Table 1: Anonymised case level extraction of all UK cases where the reported medical device is a implantable cardiac pacemaker (GMDN/CT codes 12913, 47264, 47265, 47266, 47267, 60789, 63188,CT1684, CT1685, CT257, CT2866) between 1 January 2000 up to and including 10 July 2026.

Case	First Submission Year	Admin Info Event Classification	GMDN Code	GMDN Name	GMDN CT Code	GMDN CT Name	Patient As Reported IMDRF Code	Patient As Reported IMDRF Term	Clinical Event As Reported IMDRF Code	Clinical Event As Reported IMDRF Term	Device Explanted From Date	Device Explanted To Date
1	2004	All other reportable incidents			CT257	Implantable cardiac pacemakers						
2	2004	All other reportable incidents			CT257	Implantable cardiac pacemakers						
3	2004	All other reportable incidents			CT257	Implantable cardiac pacemakers						
4	2004	All other reportable incidents			CT257	Implantable cardiac pacemakers	F1905	Device Revision or Replacement				
5	2001				CT257	Implantable cardiac pacemakers						
6	2004	All other reportable incidents			CT257	Implantable cardiac pacemakers	F1905	Device Revision or Replacement				
7	2004	All other reportable incidents			CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions				
8	2004	All other reportable incidents			CT257	Implantable cardiac pacemakers						
9	2001				CT257	Implantable cardiac pacemakers						
10	2004	All other reportable incidents			CT257	Implantable cardiac pacemakers						
11	2002				CT257	Implantable cardiac pacemakers						
12	2004	All other reportable incidents			CT257	Implantable cardiac pacemakers						
13	2002	All other reportable incidents			CT257	Implantable cardiac pacemakers						
14	2004	All other reportable incidents			CT257	Implantable cardiac pacemakers						
15	2004	All other reportable incidents			CT257	Implantable cardiac pacemakers						
16	2004	All other reportable incidents			CT257	Implantable cardiac pacemakers						
17	2004	All other reportable incidents			CT257	Implantable cardiac pacemakers						
18	2002				CT257	Implantable cardiac pacemakers						
19	2002				CT257	Implantable cardiac pacemakers						
20	2004	All other reportable incidents			CT257	Implantable cardiac pacemakers	F1905	Device Revision or Replacement				
21	2002				CT257	Implantable cardiac pacemakers						
22	2002	All other reportable incidents			CT257	Implantable cardiac pacemakers						
23	2004	All other reportable incidents			CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions				
24	2002				CT257	Implantable cardiac pacemakers						
25	2002	All other reportable incidents			CT257	Implantable cardiac pacemakers						
26	2002				CT257	Implantable cardiac pacemakers						
27	2002				CT257	Implantable cardiac pacemakers						
28	2004	All other reportable incidents			CT257	Implantable cardiac pacemakers						
29	2004	All other reportable incidents			CT257	Implantable cardiac pacemakers						
30	2002	All other reportable incidents			CT257	Implantable cardiac pacemakers						
31	2004	All other reportable incidents			CT257	Implantable cardiac pacemakers	F1905	Device Revision or Replacement				
32	2002	All other reportable incidents			CT257	Implantable cardiac pacemakers						
33	2002				CT257	Implantable cardiac pacemakers						
34	2002				CT257	Implantable cardiac pacemakers						
35	2004	All other reportable incidents			CT257	Implantable cardiac pacemakers						
36	2002	All other reportable incidents			CT257	Implantable cardiac pacemakers						
37	2002	All other reportable incidents			CT257	Implantable cardiac pacemakers						
38	2002				CT257	Implantable cardiac pacemakers						
39	2002				CT257	Implantable cardiac pacemakers						
40	2002	All other reportable incidents			CT257	Implantable cardiac pacemakers						
41	2002	All other reportable incidents			CT257	Implantable cardiac pacemakers						
42	2002	All other reportable incidents			CT257	Implantable cardiac pacemakers						
43	2002				CT257	Implantable cardiac pacemakers						
44	2004	All other reportable incidents			CT257	Implantable cardiac pacemakers						
45	2004	All other reportable incidents			CT257	Implantable cardiac pacemakers	F1905	Device Revision or Replacement				
46	2002	All other reportable incidents			CT257	Implantable cardiac pacemakers						
47	2004	All other reportable incidents			CT257	Implantable cardiac pacemakers						
48	2002				CT257	Implantable cardiac pacemakers						
49	2002				CT257	Implantable cardiac pacemakers						
50	2002				CT257	Implantable cardiac pacemakers						
51	2002				CT257	Implantable cardiac pacemakers						
52	2002				CT257	Implantable cardiac pacemakers						
53	2002	All other reportable incidents			CT257	Implantable cardiac pacemakers						
54	2002	All other reportable incidents			CT257	Implantable cardiac pacemakers						
55	2002				CT257	Implantable cardiac pacemakers						
56	2004	All other reportable incidents			CT257	Implantable cardiac pacemakers						
57	2004	All other reportable incidents			CT257	Implantable cardiac pacemakers						
58	2004	All other reportable incidents			CT257	Implantable cardiac pacemakers						
59	2004	All other reportable incidents			CT257	Implantable cardiac pacemakers						
60	2002	All other reportable incidents			CT257	Implantable cardiac pacemakers						
61	2002				CT257	Implantable cardiac pacemakers						
62	2004	All other reportable incidents			CT257	Implantable cardiac pacemakers						
63	2002				CT257	Implantable cardiac pacemakers						
64	2002	All other reportable incidents			CT257	Implantable cardiac pacemakers						
65	2002	All other reportable incidents			CT257	Implantable cardiac pacemakers						
66	2002	All other reportable incidents			CT257	Implantable cardiac pacemakers						
67	2004				CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions				
68	2002				CT257	Implantable cardiac pacemakers						
69	2002				CT257	Implantable cardiac pacemakers						
70	2002				CT257	Implantable cardiac pacemakers						
71	2004				CT257	Implantable cardiac pacemakers	F1905	Device Revision or Replacement				
72	2002				CT257	Implantable cardiac pacemakers						
73	2004				CT257	Implantable cardiac pacemakers	F1905	Device Revision or Replacement				
74	2002				CT257	Implantable cardiac pacemakers						
75	2004				CT257	Implantable cardiac pacemakers	F1905	Device Revision or Replacement				
76	2004				CT257	Implantable cardiac pacemakers						
77	2002				CT257	Implantable cardiac pacemakers						
78	2002				CT257	Implantable cardiac pacemakers						
79	2002	All other reportable incidents			CT257	Implantable cardiac pacemakers	F1905	Device Revision or Replacement				
80	2002	All other reportable incidents			CT257	Implantable cardiac pacemakers						
81	2004				CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions				
82	2004				CT257	Implantable cardiac pacemakers						

83	2002	All other reportable incidents			CT257	Implantable cardiac pacemakers						
84	2002	All other reportable incidents			CT257	Implantable cardiac pacemakers						
85	2002	All other reportable incidents			CT257	Implantable cardiac pacemakers						
86	2002	All other reportable incidents			CT257	Implantable cardiac pacemakers						
87	2004				CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions				
88	2002				CT257	Implantable cardiac pacemakers						
89	2002				CT257	Implantable cardiac pacemakers						
90	2004				CT257	Implantable cardiac pacemakers						
91	2002				CT257	Implantable cardiac pacemakers						
92	2002				CT257	Implantable cardiac pacemakers						
93	2002				CT257	Implantable cardiac pacemakers						
94	2004				CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions				
95	2002	All other reportable incidents			CT257	Implantable cardiac pacemakers						
96	2002				CT257	Implantable cardiac pacemakers						
97	2002				CT257	Implantable cardiac pacemakers						
98	2002	All other reportable incidents			CT257	Implantable cardiac pacemakers						
99	2004				CT257	Implantable cardiac pacemakers	F1905	Device Revision or Replacement				
100	2002				CT257	Implantable cardiac pacemakers						
101	2002	All other reportable incidents			CT257	Implantable cardiac pacemakers						
102	2002	All other reportable incidents			CT257	Implantable cardiac pacemakers						
103	2002	All other reportable incidents			CT257	Implantable cardiac pacemakers						
104	2002	All other reportable incidents			CT257	Implantable cardiac pacemakers						
105	2002	All other reportable incidents			CT257	Implantable cardiac pacemakers						
106	2004				CT257	Implantable cardiac pacemakers						
107	2004				CT257	Implantable cardiac pacemakers	F1905	Device Revision or Replacement				
108	2002				CT257	Implantable cardiac pacemakers						
109	2002				CT257	Implantable cardiac pacemakers						
110	2002	All other reportable incidents			CT257	Implantable cardiac pacemakers						
111	2004				CT257	Implantable cardiac pacemakers	F1905	Device Revision or Replacement				
112	2004				CT257	Implantable cardiac pacemakers	F1905	Device Revision or Replacement				
113	2004				CT257	Implantable cardiac pacemakers	F1905	Device Revision or Replacement				
114	2002				CT257	Implantable cardiac pacemakers						
115	2002				CT257	Implantable cardiac pacemakers						
116	2000				CT257	Implantable cardiac pacemakers						
117	2000				CT257	Implantable cardiac pacemakers						
118	2002				CT257	Implantable cardiac pacemakers						
119	2002	All other reportable incidents			CT257	Implantable cardiac pacemakers						
120	2000				CT257	Implantable cardiac pacemakers						
121	2000				CT257	Implantable cardiac pacemakers						
122	2002				CT257	Implantable cardiac pacemakers						
123	2002				CT257	Implantable cardiac pacemakers						
124	2000				CT257	Implantable cardiac pacemakers						
125	2000				CT257	Implantable cardiac pacemakers						
126	2004				CT257	Implantable cardiac pacemakers	F1905	Device Revision or Replacement				
127	2000		45966	Transthoracic pacing lead	CT1684	Cardiac pacemakers and associated devices						
128	2002				CT257	Implantable cardiac pacemakers						
129	2000				CT257	Implantable cardiac pacemakers						
130	2000				CT257	Implantable cardiac pacemakers						
131	2002	All other reportable incidents			CT257	Implantable cardiac pacemakers						
132	2002	All other reportable incidents			CT257	Implantable cardiac pacemakers						
133	2002	All other reportable incidents			CT257	Implantable cardiac pacemakers						
134	2000				CT257	Implantable cardiac pacemakers						
135	2002				CT257	Implantable cardiac pacemakers						
136	2000				CT257	Implantable cardiac pacemakers						
137	2000				CT257	Implantable cardiac pacemakers						
138	2000				CT257	Implantable cardiac pacemakers						
139	2004				CT257	Implantable cardiac pacemakers	F1905	Device Revision or Replacement				
140	2000	All other reportable incidents			CT257	Implantable cardiac pacemakers						
141	2000				CT257	Implantable cardiac pacemakers						
142	2000				CT257	Implantable cardiac pacemakers						
143	2002	All other reportable incidents			CT257	Implantable cardiac pacemakers						
144	2004				CT257	Implantable cardiac pacemakers	F1905	Device Revision or Replacement				
145	2000				CT257	Implantable cardiac pacemakers						
146	2002	All other reportable incidents			CT257	Implantable cardiac pacemakers						
147	2000				CT257	Implantable cardiac pacemakers						
148	2004				CT257	Implantable cardiac pacemakers	F1905	Device Revision or Replacement				
149	2000				CT257	Implantable cardiac pacemakers						
150	2000				CT257	Implantable cardiac pacemakers						
151	2000				CT257	Implantable cardiac pacemakers						
152	2004				CT257	Implantable cardiac pacemakers	F1905	Device Revision or Replacement				
153	2004				CT257	Implantable cardiac pacemakers	F1905	Device Revision or Replacement				
154	2000				CT257	Implantable cardiac pacemakers						
155	2002	All other reportable incidents			CT257	Implantable cardiac pacemakers						
156	2000				CT257	Implantable cardiac pacemakers						
157	2004				CT257	Implantable cardiac pacemakers	F1905	Device Revision or Replacement				
158	2002				CT257	Implantable cardiac pacemakers						
159	2002	All other reportable incidents			CT257	Implantable cardiac pacemakers						
160	2004				CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions				
161	2002				CT257	Implantable cardiac pacemakers						
162	2002				CT257	Implantable cardiac pacemakers						
163	2004				CT257	Implantable cardiac pacemakers	F1905	Device Revision or Replacement				
164	2000				CT257	Implantable cardiac pacemakers						
165	2002				CT257	Implantable cardiac pacemakers						
166	2000				CT257	Implantable cardiac pacemakers						
167	2000				CT257	Implantable cardiac pacemakers						
168	2000				CT257	Implantable cardiac pacemakers						
169	2002	All other reportable incidents			CT257	Implantable cardiac pacemakers						
170	2000				CT257	Implantable cardiac pacemakers						

171	2002				CT257	Implantable cardiac pacemakers							
172	2004				CT257	Implantable cardiac pacemakers	F1905	Device Revision or Replacement					
173	2002	All other reportable incidents			CT257	Implantable cardiac pacemakers							
174	2002				CT257	Implantable cardiac pacemakers							
175	2002	All other reportable incidents			CT257	Implantable cardiac pacemakers							
176	2000				CT257	Implantable cardiac pacemakers							
177	2004				CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions					
178	2004				CT257	Implantable cardiac pacemakers	F1905	Device Revision or Replacement					
179	2004				CT257	Implantable cardiac pacemakers	F1905	Device Revision or Replacement					
180	2002	All other reportable incidents			CT257	Implantable cardiac pacemakers							
181	2002	All other reportable incidents			CT257	Implantable cardiac pacemakers							
182	2000				CT257	Implantable cardiac pacemakers							
183	2002	All other reportable incidents			CT257	Implantable cardiac pacemakers							
184	2002				CT257	Implantable cardiac pacemakers							
185	2004				CT257	Implantable cardiac pacemakers	F1905	Device Revision or Replacement					
186	2002	All other reportable incidents			CT257	Implantable cardiac pacemakers							
187	2002	All other reportable incidents			CT257	Implantable cardiac pacemakers							
188	2000				CT257	Implantable cardiac pacemakers							
189	2004				CT257	Implantable cardiac pacemakers	F1905	Device Revision or Replacement					
190	2002	All other reportable incidents			CT257	Implantable cardiac pacemakers							
191	2000				CT257	Implantable cardiac pacemakers							
192	2000				CT257	Implantable cardiac pacemakers							
193	2000	All other reportable incidents			CT257	Implantable cardiac pacemakers							
194	2002	All other reportable incidents			CT257	Implantable cardiac pacemakers							
195	2004				CT257	Implantable cardiac pacemakers							
196	2000	All other reportable incidents			CT257	Implantable cardiac pacemakers							
197	2002				CT257	Implantable cardiac pacemakers							
198	2000				CT257	Implantable cardiac pacemakers							
199	2000				CT257	Implantable cardiac pacemakers							
200	2000				CT257	Implantable cardiac pacemakers							
201	2002	All other reportable incidents			CT257	Implantable cardiac pacemakers							
202	2004				CT257	Implantable cardiac pacemakers	F1905	Device Revision or Replacement					
203	2002	All other reportable incidents			CT257	Implantable cardiac pacemakers							
204	2000				CT257	Implantable cardiac pacemakers							
205	2002	All other reportable incidents			CT257	Implantable cardiac pacemakers							
206	2000				CT257	Implantable cardiac pacemakers							
207	2002				CT257	Implantable cardiac pacemakers							
208	2004				CT257	Implantable cardiac pacemakers	F1905	Device Revision or Replacement					
209	2002	All other reportable incidents			CT257	Implantable cardiac pacemakers							
210	2000				CT257	Implantable cardiac pacemakers							
211	2002	All other reportable incidents			CT257	Implantable cardiac pacemakers							
212	2000				CT257	Implantable cardiac pacemakers							
213	2000				CT257	Implantable cardiac pacemakers							
214	2000				CT257	Implantable cardiac pacemakers							
215	2002				CT257	Implantable cardiac pacemakers							
216	2000				CT257	Implantable cardiac pacemakers							
217	2000				CT257	Implantable cardiac pacemakers							
218	2000				CT257	Implantable cardiac pacemakers							
219	2000				CT257	Implantable cardiac pacemakers							
220	2002				CT257	Implantable cardiac pacemakers							
221	2000				CT257	Implantable cardiac pacemakers							
222	2000				CT257	Implantable cardiac pacemakers							
223	2000				CT257	Implantable cardiac pacemakers							
224	2000				CT257	Implantable cardiac pacemakers							
225	2000				CT257	Implantable cardiac pacemakers							
226	2000				CT257	Implantable cardiac pacemakers							
227	2002	All other reportable incidents			CT257	Implantable cardiac pacemakers							
228	2000				CT257	Implantable cardiac pacemakers							
229	2000				CT257	Implantable cardiac pacemakers							
230	2002	All other reportable incidents			CT257	Implantable cardiac pacemakers							
231	2000				CT257	Implantable cardiac pacemakers							
232	2004				CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions					
233	2000				CT257	Implantable cardiac pacemakers							
234	2002				CT257	Implantable cardiac pacemakers							
235	2000				CT257	Implantable cardiac pacemakers							
236	2000				CT257	Implantable cardiac pacemakers							
237	2000				CT257	Implantable cardiac pacemakers							
238	2000	All other reportable incidents			CT257	Implantable cardiac pacemakers							
239	2002				CT257	Implantable cardiac pacemakers							
240	2000				CT257	Implantable cardiac pacemakers							
241	2000				CT257	Implantable cardiac pacemakers							
242	2004				CT257	Implantable cardiac pacemakers	F1905	Device Revision or Replacement					
243	2000				CT257	Implantable cardiac pacemakers							
244	2002	All other reportable incidents			CT257	Implantable cardiac pacemakers							
245	2002				CT257	Implantable cardiac pacemakers							
246	2000				CT257	Implantable cardiac pacemakers							
247	2000				CT257	Implantable cardiac pacemakers							
248	2000				CT257	Implantable cardiac pacemakers							
249	2000				CT257	Implantable cardiac pacemakers							
250	2000				CT257	Implantable cardiac pacemakers							
251	2000				CT257	Implantable cardiac pacemakers							
252	2002				CT257	Implantable cardiac pacemakers							
253	2002	All other reportable incidents			CT257	Implantable cardiac pacemakers							
254	2004				CT257	Implantable cardiac pacemakers	F1905	Device Revision or Replacement					
255	2002	All other reportable incidents			CT257	Implantable cardiac pacemakers							
256	2002	All other reportable incidents			CT257	Implantable cardiac pacemakers							
257	2002	All other reportable incidents			CT257	Implantable cardiac pacemakers							
258	2002	All other reportable incidents			CT257	Implantable cardiac pacemakers							

259	2002				CT257	Implantable cardiac pacemakers						
260	2002				CT257	Implantable cardiac pacemakers						
261	2002	All other reportable incidents			CT257	Implantable cardiac pacemakers						
262	2002	All other reportable incidents			CT257	Implantable cardiac pacemakers						
263	2000				CT257	Implantable cardiac pacemakers						
264	2004				CT257	Implantable cardiac pacemakers	F1905	Device Revision or Replacement				
265	2004				CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions				
266	2002				CT257	Implantable cardiac pacemakers						
267	2004				CT257	Implantable cardiac pacemakers						
268	2000				CT257	Implantable cardiac pacemakers						
269	2002				CT257	Implantable cardiac pacemakers						
270	2004				CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions				
271	2002				CT257	Implantable cardiac pacemakers						
272	2004	All other reportable incidents			CT257	Implantable cardiac pacemakers						
273	2004				CT257	Implantable cardiac pacemakers	F1905	Device Revision or Replacement				
274	2000				CT257	Implantable cardiac pacemakers						
275	2000				CT257	Implantable cardiac pacemakers						
276	2000				CT257	Implantable cardiac pacemakers						
277	2002	All other reportable incidents			CT257	Implantable cardiac pacemakers						
278	2002	All other reportable incidents			CT257	Implantable cardiac pacemakers						
279	2000				CT257	Implantable cardiac pacemakers						
280	2000				CT257	Implantable cardiac pacemakers						
281	2000				CT257	Implantable cardiac pacemakers						
282	2000				CT257	Implantable cardiac pacemakers						
283	2000				CT257	Implantable cardiac pacemakers						
284	2000				CT257	Implantable cardiac pacemakers						
285	2002	All other reportable incidents			CT257	Implantable cardiac pacemakers						
286	2002	All other reportable incidents			CT257	Implantable cardiac pacemakers						
287	2005				CT257	Implantable cardiac pacemakers	F1905	Device Revision or Replacement				
288	2000				CT257	Implantable cardiac pacemakers						
289	2002				CT257	Implantable cardiac pacemakers						
290	2005				CT257	Implantable cardiac pacemakers						
291	2000	All other reportable incidents			CT257	Implantable cardiac pacemakers						
292	2002	All other reportable incidents			CT257	Implantable cardiac pacemakers						
293	2005				CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions				
294	2000				CT257	Implantable cardiac pacemakers						
295	2000				CT257	Implantable cardiac pacemakers						
296	2000				CT257	Implantable cardiac pacemakers						
297	2000				CT257	Implantable cardiac pacemakers						
298	2000				CT257	Implantable cardiac pacemakers						
299	2000				CT257	Implantable cardiac pacemakers						
300	2000				CT257	Implantable cardiac pacemakers						
301	2005				CT257	Implantable cardiac pacemakers	F1905	Device Revision or Replacement				
302	2000				CT257	Implantable cardiac pacemakers						
303	2000				CT257	Implantable cardiac pacemakers						
304	2000				CT257	Implantable cardiac pacemakers						
305	2005				CT257	Implantable cardiac pacemakers	F1905	Device Revision or Replacement				
306	2000				CT257	Implantable cardiac pacemakers						
307	2002	All other reportable incidents			CT257	Implantable cardiac pacemakers						
308	2002				CT257	Implantable cardiac pacemakers						
309	2005				CT257	Implantable cardiac pacemakers	F1905	Device Revision or Replacement				
310	2005				CT257	Implantable cardiac pacemakers	F1905	Device Revision or Replacement				
311	2002	All other reportable incidents			CT257	Implantable cardiac pacemakers						
312	2002	All other reportable incidents			CT257	Implantable cardiac pacemakers						
313	2002				CT257	Implantable cardiac pacemakers						
314	2002				CT257	Implantable cardiac pacemakers						
315	2002	All other reportable incidents			CT257	Implantable cardiac pacemakers						
316	2000	All other reportable incidents			CT257	Implantable cardiac pacemakers						
317	2002	All other reportable incidents			CT257	Implantable cardiac pacemakers						
318	2002	All other reportable incidents			CT257	Implantable cardiac pacemakers						
319	2002	All other reportable incidents			CT257	Implantable cardiac pacemakers						
320	2000				CT257	Implantable cardiac pacemakers						
321	2002	All other reportable incidents			CT257	Implantable cardiac pacemakers						
322	2005				CT257	Implantable cardiac pacemakers	F1905	Device Revision or Replacement				
323	2000				CT257	Implantable cardiac pacemakers						
324	2005				CT257	Implantable cardiac pacemakers	F1905	Device Revision or Replacement				
325	2002	All other reportable incidents			CT257	Implantable cardiac pacemakers						
326	2000				CT257	Implantable cardiac pacemakers						
327	2002	All other reportable incidents	45966	Transthoracic pacing lead	CT1684	Cardiac pacemakers and associated devices						
328	2002				CT257	Implantable cardiac pacemakers						
329	2005				CT257	Implantable cardiac pacemakers						
330	2005				CT257	Implantable cardiac pacemakers	F1905	Device Revision or Replacement				
331	2000				CT257	Implantable cardiac pacemakers						
332	2000				CT257	Implantable cardiac pacemakers						
333	2000				CT257	Implantable cardiac pacemakers						
334	2003	All other reportable incidents			CT257	Implantable cardiac pacemakers						
335	2003	All other reportable incidents			CT257	Implantable cardiac pacemakers						
336	2000	All other reportable incidents			CT257	Implantable cardiac pacemakers						
337	2005				CT257	Implantable cardiac pacemakers	F1905	Device Revision or Replacement				
338	2003	All other reportable incidents			CT257	Implantable cardiac pacemakers						
339	2003	All other reportable incidents			CT257	Implantable cardiac pacemakers						
340	2000				CT257	Implantable cardiac pacemakers						
341	2000				CT257	Implantable cardiac pacemakers						
342	2000				CT257	Implantable cardiac pacemakers						
343	2000				CT257	Implantable cardiac pacemakers						
344	2005				CT257	Implantable cardiac pacemakers	F1905	Device Revision or Replacement				
345	2005				CT257	Implantable cardiac pacemakers	F1905	Device Revision or Replacement				
346	2000				CT257	Implantable cardiac pacemakers						

347	2003	All other reportable incidents			CT257	Implantable cardiac pacemakers	F1905	Device Revision or Replacement				
348	2003	All other reportable incidents			CT257	Implantable cardiac pacemakers						
349	2003	All other reportable incidents			CT257	Implantable cardiac pacemakers						
350	2000				CT257	Implantable cardiac pacemakers						
351	2000				CT257	Implantable cardiac pacemakers						
352	2003	All other reportable incidents			CT257	Implantable cardiac pacemakers						
353	2003	All other reportable incidents			CT257	Implantable cardiac pacemakers						
354	2003	All other reportable incidents			CT257	Implantable cardiac pacemakers	F1905	Device Revision or Replacement				
355	2003	All other reportable incidents			CT257	Implantable cardiac pacemakers						
356	2003	All other reportable incidents			CT257	Implantable cardiac pacemakers						
357	2005				CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions				
358	2005				CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions				
359	2000				CT257	Implantable cardiac pacemakers						
360	2003	All other reportable incidents			CT257	Implantable cardiac pacemakers						
361	2005				CT257	Implantable cardiac pacemakers	F1905	Device Revision or Replacement				
362	2000				CT257	Implantable cardiac pacemakers						
363	2000				CT257	Implantable cardiac pacemakers						
364	2003	All other reportable incidents			CT257	Implantable cardiac pacemakers						
365	2000				CT257	Implantable cardiac pacemakers						
366	2003	All other reportable incidents			CT257	Implantable cardiac pacemakers						
367	2000				CT257	Implantable cardiac pacemakers						
368	2000				CT257	Implantable cardiac pacemakers						
369	2000				CT257	Implantable cardiac pacemakers						
370	2003	All other reportable incidents			CT257	Implantable cardiac pacemakers						
371	2003	All other reportable incidents			CT257	Implantable cardiac pacemakers						
372	2005				CT257	Implantable cardiac pacemakers	F1905	Device Revision or Replacement				
373	2000				CT257	Implantable cardiac pacemakers						
374	2000	All other reportable incidents			CT257	Implantable cardiac pacemakers						
375	2000				CT257	Implantable cardiac pacemakers						
376	2005				CT257	Implantable cardiac pacemakers	F1905	Device Revision or Replacement				
377	2003	All other reportable incidents			CT257	Implantable cardiac pacemakers						
378	2005				CT257	Implantable cardiac pacemakers	F1905	Device Revision or Replacement				
379	2005				CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions				
380	2000				CT257	Implantable cardiac pacemakers						
381	2000				CT257	Implantable cardiac pacemakers						
382	2003	All other reportable incidents			CT257	Implantable cardiac pacemakers						
383	2003	All other reportable incidents			CT257	Implantable cardiac pacemakers						
384	2003	All other reportable incidents			CT257	Implantable cardiac pacemakers						
385	2005				CT257	Implantable cardiac pacemakers	F1905	Device Revision or Replacement				
386	2000				CT257	Implantable cardiac pacemakers						
387	2005				CT257	Implantable cardiac pacemakers	F1905	Device Revision or Replacement				
388	2003	All other reportable incidents			CT257	Implantable cardiac pacemakers						
389	2003	All other reportable incidents			CT257	Implantable cardiac pacemakers						
390	2005		60910	Coronary venous pacing lead	CT1684	Cardiac pacemakers and associated devices	F1905	Device Revision or Replacement				
391	2005				CT257	Implantable cardiac pacemakers	F1905	Device Revision or Replacement				
392	2005				CT257	Implantable cardiac pacemakers	F1905	Device Revision or Replacement				
393	2000				CT257	Implantable cardiac pacemakers						
394	2005				CT257	Implantable cardiac pacemakers	F1905	Device Revision or Replacement				
395	2000				CT257	Implantable cardiac pacemakers						
396	2000				CT257	Implantable cardiac pacemakers						
397	2005				CT257	Implantable cardiac pacemakers	F1905	Device Revision or Replacement				
398	2003	All other reportable incidents			CT257	Implantable cardiac pacemakers						
399	2000				CT257	Implantable cardiac pacemakers						
400	2000	All other reportable incidents			CT257	Implantable cardiac pacemakers						
401	2000	All other reportable incidents			CT257	Implantable cardiac pacemakers						
402	2000	All other reportable incidents			CT257	Implantable cardiac pacemakers						
403	2003	All other reportable incidents			CT257	Implantable cardiac pacemakers						
404	2005				CT257	Implantable cardiac pacemakers	F1905	Device Revision or Replacement				
405	2000				CT257	Implantable cardiac pacemakers						
406	2000	All other reportable incidents			CT257	Implantable cardiac pacemakers						
407	2003	All other reportable incidents			CT257	Implantable cardiac pacemakers						
408	2000				CT257	Implantable cardiac pacemakers						
409	2000				CT257	Implantable cardiac pacemakers						
410	2003	All other reportable incidents			CT257	Implantable cardiac pacemakers						
411	2003	All other reportable incidents			CT257	Implantable cardiac pacemakers						
412	2003	All other reportable incidents			CT257	Implantable cardiac pacemakers						
413	2005				CT257	Implantable cardiac pacemakers	F1905	Device Revision or Replacement				
414	2005				CT257	Implantable cardiac pacemakers	F1905	Device Revision or Replacement				
415	2005				CT257	Implantable cardiac pacemakers	F1905	Device Revision or Replacement				
416	2005				CT257	Implantable cardiac pacemakers	F1905	Device Revision or Replacement				
417	2005		60910	Coronary venous pacing lead	CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions				
418	2000				CT257	Implantable cardiac pacemakers						
419	2003	All other reportable incidents			CT257	Implantable cardiac pacemakers						
420	2005				CT257	Implantable cardiac pacemakers	F1905	Device Revision or Replacement				
421	2003	All other reportable incidents			CT257	Implantable cardiac pacemakers						
422	2005				CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions				
423	2003	All other reportable incidents			CT257	Implantable cardiac pacemakers						
424	2003	All other reportable incidents			CT257	Implantable cardiac pacemakers						
425	2005				CT257	Implantable cardiac pacemakers	F1905	Device Revision or Replacement				
426	2005				CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions				
427	2005				CT257	Implantable cardiac pacemakers	F1905	Device Revision or Replacement				
428	2005				CT257	Implantable cardiac pacemakers	F1905	Device Revision or Replacement				
429	2003	All other reportable incidents			CT257	Implantable cardiac pacemakers						
430	2000				CT257	Implantable cardiac pacemakers						
431	2000				CT257	Implantable cardiac pacemakers						
432	2003	All other reportable incidents			CT257	Implantable cardiac pacemakers						
433	2005				CT257	Implantable cardiac pacemakers	F1905	Device Revision or Replacement				
434	2003	All other reportable incidents			CT257	Implantable cardiac pacemakers	F1905	Device Revision or Replacement				

435	2003	All other reportable incidents			CT257	Implantable cardiac pacemakers						
436	2005				CT257	Implantable cardiac pacemakers	F1905	Device Revision or Replacement				
437	2005				CT257	Implantable cardiac pacemakers	F1905	Device Revision or Replacement				
438	2000	All other reportable incidents			CT257	Implantable cardiac pacemakers						
439	2005				CT257	Implantable cardiac pacemakers	F1905	Device Revision or Replacement				
440	2000				CT257	Implantable cardiac pacemakers						
441	2003	All other reportable incidents			CT257	Implantable cardiac pacemakers						
442	2000				CT257	Implantable cardiac pacemakers						
443	2005				CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions				
444	2005				CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions				
445	2000				CT257	Implantable cardiac pacemakers						
446	2000				CT257	Implantable cardiac pacemakers						
447	2005				CT257	Implantable cardiac pacemakers	F1905	Device Revision or Replacement				
448	2003	All other reportable incidents			CT257	Implantable cardiac pacemakers						
449	2003	All other reportable incidents			CT257	Implantable cardiac pacemakers						
450	2000				CT257	Implantable cardiac pacemakers						
451	2000				CT257	Implantable cardiac pacemakers						
452	2005				CT257	Implantable cardiac pacemakers	F1905	Device Revision or Replacement				
453	2000				CT257	Implantable cardiac pacemakers						
454	2003	All other reportable incidents			CT257	Implantable cardiac pacemakers						
455	2003	All other reportable incidents			CT257	Implantable cardiac pacemakers						
456	2005				CT257	Implantable cardiac pacemakers	F1905	Device Revision or Replacement				
457	2003	All other reportable incidents			CT257	Implantable cardiac pacemakers						
458	2003	All other reportable incidents			CT257	Implantable cardiac pacemakers						
459	2000				CT257	Implantable cardiac pacemakers						
460	2000				CT257	Implantable cardiac pacemakers						
461	2000				CT257	Implantable cardiac pacemakers						
462	2000				CT257	Implantable cardiac pacemakers						
463	2005				CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions				
464	2005				CT257	Implantable cardiac pacemakers	F1905	Device Revision or Replacement				
465	2000				CT257	Implantable cardiac pacemakers						
466	2000				CT257	Implantable cardiac pacemakers						
467	2000				CT257	Implantable cardiac pacemakers						
468	2000				CT257	Implantable cardiac pacemakers						
469	2000				CT257	Implantable cardiac pacemakers						
470	2000				CT257	Implantable cardiac pacemakers						
471	2003	All other reportable incidents			CT257	Implantable cardiac pacemakers						
472	2000				CT257	Implantable cardiac pacemakers						
473	2003	All other reportable incidents	45966	Transthoracic pacing lead	CT1684	Cardiac pacemakers and associated devices						
474	2000				CT257	Implantable cardiac pacemakers						
475	2000				CT257	Implantable cardiac pacemakers						
476	2000				CT257	Implantable cardiac pacemakers						
477	2003	All other reportable incidents			CT257	Implantable cardiac pacemakers						
478	2000				CT257	Implantable cardiac pacemakers						
479	2003	All other reportable incidents			CT257	Implantable cardiac pacemakers						
480	2000	All other reportable incidents			CT257	Implantable cardiac pacemakers						
481	2000				CT257	Implantable cardiac pacemakers						
482	2005				CT257	Implantable cardiac pacemakers	F1905	Device Revision or Replacement				
483	2000				CT257	Implantable cardiac pacemakers						
484	2003	All other reportable incidents			CT257	Implantable cardiac pacemakers						
485	2003	All other reportable incidents			CT257	Implantable cardiac pacemakers						
486	2003	All other reportable incidents			CT257	Implantable cardiac pacemakers						
487	2000				CT257	Implantable cardiac pacemakers						
488	2003	All other reportable incidents			CT257	Implantable cardiac pacemakers						
489	2000				CT257	Implantable cardiac pacemakers						
490	2005		47270	Cardiac resynchronization therapy implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions				
491	2000	All other reportable incidents			CT257	Implantable cardiac pacemakers						
492	2005				CT257	Implantable cardiac pacemakers	F1905	Device Revision or Replacement				
493	2000				CT257	Implantable cardiac pacemakers						
494	2000				CT257	Implantable cardiac pacemakers						
495	2000				CT257	Implantable cardiac pacemakers						
496	2000				CT257	Implantable cardiac pacemakers						
497	2005				CT257	Implantable cardiac pacemakers	F1905	Device Revision or Replacement				
498	2005				CT257	Implantable cardiac pacemakers	F1905	Device Revision or Replacement				
499	2000	All other reportable incidents			CT257	Implantable cardiac pacemakers						
500	2000				CT257	Implantable cardiac pacemakers						
501	2000				CT257	Implantable cardiac pacemakers						
502	2000				CT257	Implantable cardiac pacemakers						
503	2003	All other reportable incidents			CT257	Implantable cardiac pacemakers						
504	2000				CT257	Implantable cardiac pacemakers						
505	2003	All other reportable incidents			CT257	Implantable cardiac pacemakers						
506	2005				CT257	Implantable cardiac pacemakers	F1905	Device Revision or Replacement				
507	2003	All other reportable incidents			CT257	Implantable cardiac pacemakers						
508	2003	All other reportable incidents			CT257	Implantable cardiac pacemakers						
509	2005				CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions				
510	2000				CT257	Implantable cardiac pacemakers						
511	2000				CT257	Implantable cardiac pacemakers						
512	2003	All other reportable incidents			CT257	Implantable cardiac pacemakers						
513	2003	All other reportable incidents			CT257	Implantable cardiac pacemakers						
514	2003	All other reportable incidents			CT257	Implantable cardiac pacemakers						
515	2005				CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions				
516	2003	All other reportable incidents			CT257	Implantable cardiac pacemakers						
517	2003	All other reportable incidents			CT257	Implantable cardiac pacemakers						
518	2003	All other reportable incidents			CT257	Implantable cardiac pacemakers						
519	2001				CT257	Implantable cardiac pacemakers						
520	2001				CT257	Implantable cardiac pacemakers						
521	2003	All other reportable incidents			CT257	Implantable cardiac pacemakers						

522	2003	All other reportable incidents			CT257	Implantable cardiac pacemakers						
523	2001				CT257	Implantable cardiac pacemakers						
524	2003	All other reportable incidents			CT257	Implantable cardiac pacemakers						
525	2003	All other reportable incidents			CT257	Implantable cardiac pacemakers						
526	2005				CT257	Implantable cardiac pacemakers	F1905	Device Revision or Replacement				
527	2005				CT257	Implantable cardiac pacemakers	F1905	Device Revision or Replacement				
528	2005		47270	Cardiac resynchronization therapy implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions				
529	2001	All other reportable incidents			CT257	Implantable cardiac pacemakers						
530	2003	All other reportable incidents			CT257	Implantable cardiac pacemakers	F1905	Device Revision or Replacement				
531	2003	All other reportable incidents			CT257	Implantable cardiac pacemakers						
532	2001				CT257	Implantable cardiac pacemakers						
533	2003	All other reportable incidents			CT257	Implantable cardiac pacemakers						
534	2003	All other reportable incidents			CT257	Implantable cardiac pacemakers						
535	2001	All other reportable incidents			CT257	Implantable cardiac pacemakers						
536	2001				CT257	Implantable cardiac pacemakers						
537	2005				CT257	Implantable cardiac pacemakers	F1905	Device Revision or Replacement				
538	2001				CT257	Implantable cardiac pacemakers						
539	2001				CT257	Implantable cardiac pacemakers						
540	2003	All other reportable incidents			CT257	Implantable cardiac pacemakers						
541	2005				CT257	Implantable cardiac pacemakers	F1905	Device Revision or Replacement				
542	2001				CT257	Implantable cardiac pacemakers						
543	2005				CT257	Implantable cardiac pacemakers	F1905	Device Revision or Replacement				
544	2003	All other reportable incidents			CT257	Implantable cardiac pacemakers						
545	2003				CT257	Implantable cardiac pacemakers						
546	2003	All other reportable incidents			CT257	Implantable cardiac pacemakers						
547	2003	All other reportable incidents			CT257	Implantable cardiac pacemakers	F1905	Device Revision or Replacement				
548	2001				CT257	Implantable cardiac pacemakers						
549	2003	All other reportable incidents			CT257	Implantable cardiac pacemakers	F1905	Device Revision or Replacement				
550	2001		45966	Transthoracic pacing lead	CT1684	Cardiac pacemakers and associated devices						
551	2001				CT257	Implantable cardiac pacemakers						
552	2003	All other reportable incidents			CT257	Implantable cardiac pacemakers						
553	2003	All other reportable incidents			CT257	Implantable cardiac pacemakers						
554	2005				CT257	Implantable cardiac pacemakers	F1905	Device Revision or Replacement				
555	2003	All other reportable incidents			CT257	Implantable cardiac pacemakers						
556	2001				CT257	Implantable cardiac pacemakers						
557	2003	All other reportable incidents			CT257	Implantable cardiac pacemakers						
558	2003	All other reportable incidents			CT257	Implantable cardiac pacemakers						
559	2005				CT257	Implantable cardiac pacemakers	F1905	Device Revision or Replacement				
560	2005				CT257	Implantable cardiac pacemakers	F1905	Device Revision or Replacement				
561	2005				CT257	Implantable cardiac pacemakers	F1905	Device Revision or Replacement				
562	2001	All other reportable incidents			CT257	Implantable cardiac pacemakers						
563	2005				CT257	Implantable cardiac pacemakers	F1905	Device Revision or Replacement				
564	2003	All other reportable incidents			CT257	Implantable cardiac pacemakers						
565	2001				CT257	Implantable cardiac pacemakers						
566	2003	All other reportable incidents			CT257	Implantable cardiac pacemakers						
567	2003	All other reportable incidents			CT257	Implantable cardiac pacemakers						
568	2003				CT257	Implantable cardiac pacemakers						
569	2001				CT257	Implantable cardiac pacemakers						
570	2001				CT257	Implantable cardiac pacemakers						
571	2005				CT257	Implantable cardiac pacemakers	F1905	Device Revision or Replacement				
572	2001				CT257	Implantable cardiac pacemakers						
573	2001				CT257	Implantable cardiac pacemakers						
574	2005				CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions				
575	2001				CT257	Implantable cardiac pacemakers						
576	2005				CT257	Implantable cardiac pacemakers	F1905	Device Revision or Replacement				
577	2005				CT257	Implantable cardiac pacemakers	F1905	Device Revision or Replacement				
578	2001	All other reportable incidents			CT257	Implantable cardiac pacemakers						
579	2001				CT257	Implantable cardiac pacemakers						
580	2001				CT257	Implantable cardiac pacemakers						
581	2001	All other reportable incidents			CT257	Implantable cardiac pacemakers						
582	2001	All other reportable incidents			CT257	Implantable cardiac pacemakers						
583	2003	All other reportable incidents			CT257	Implantable cardiac pacemakers						
584	2001	All other reportable incidents			CT257	Implantable cardiac pacemakers						
585	2005				CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions				
586	2001				CT257	Implantable cardiac pacemakers						
587	2005		60910	Coronary venous pacing lead	CT1684	Cardiac pacemakers and associated devices	F1905	Device Revision or Replacement				
588	2001				CT257	Implantable cardiac pacemakers						
589	2003	All other reportable incidents			CT257	Implantable cardiac pacemakers						
590	2005				CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions				
591	2001	All other reportable incidents			CT257	Implantable cardiac pacemakers						
592	2001	All other reportable incidents			CT257	Implantable cardiac pacemakers						
593	2005				CT257	Implantable cardiac pacemakers						
594	2001				CT257	Implantable cardiac pacemakers						
595	2003	All other reportable incidents			CT257	Implantable cardiac pacemakers	F1905	Device Revision or Replacement				
596	2001				CT257	Implantable cardiac pacemakers						
597	2001	All other reportable incidents			CT257	Implantable cardiac pacemakers						
598	2005				CT257	Implantable cardiac pacemakers	F1905	Device Revision or Replacement				
599	2003	All other reportable incidents			CT257	Implantable cardiac pacemakers						
600	2003	All other reportable incidents			CT257	Implantable cardiac pacemakers						
601	2001				CT257	Implantable cardiac pacemakers						
602	2005				CT257	Implantable cardiac pacemakers	F1905	Device Revision or Replacement				
603	2005				CT257	Implantable cardiac pacemakers	F1905	Device Revision or Replacement				
604	2001	All other reportable incidents			CT257	Implantable cardiac pacemakers						
605	2001				CT257	Implantable cardiac pacemakers						
606	2001				CT257	Implantable cardiac pacemakers						
607	2003				CT257	Implantable cardiac pacemakers						
608	2001				CT257	Implantable cardiac pacemakers						

609	2003	All other reportable incidents			CT257	Implantable cardiac pacemakers						
610	2001	All other reportable incidents			CT257	Implantable cardiac pacemakers						
611	2001	All other reportable incidents			CT257	Implantable cardiac pacemakers						
612	2003	All other reportable incidents			CT257	Implantable cardiac pacemakers						
613	2001	All other reportable incidents			CT257	Implantable cardiac pacemakers						
614	2001	All other reportable incidents			CT257	Implantable cardiac pacemakers						
615	2005				CT257	Implantable cardiac pacemakers	F1905	Device Revision or Replacement				
616	2005				CT257	Implantable cardiac pacemakers	F1905	Device Revision or Replacement				
617	2005				CT257	Implantable cardiac pacemakers	F1905	Device Revision or Replacement				
618	2001				CT257	Implantable cardiac pacemakers						
619	2005				CT257	Implantable cardiac pacemakers	F1905	Device Revision or Replacement				
620	2005				CT257	Implantable cardiac pacemakers	F1905	Device Revision or Replacement				
621	2005				CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions				
622	2005				CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions				
623	2005				CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions				
624	2005				CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions				
625	2001				CT257	Implantable cardiac pacemakers						
626	2001	All other reportable incidents			CT257	Implantable cardiac pacemakers						
627	2005				CT257	Implantable cardiac pacemakers	F1905	Device Revision or Replacement				
628	2001		45966	Transthoracic pacing lead	CT1684	Cardiac pacemakers and associated devices						
629	2005				CT257	Implantable cardiac pacemakers	F1905	Device Revision or Replacement				
630	2001				CT257	Implantable cardiac pacemakers						
631	2005				CT257	Implantable cardiac pacemakers	F1905	Device Revision or Replacement				
632	2001				CT257	Implantable cardiac pacemakers						
633	2003	All other reportable incidents			CT257	Implantable cardiac pacemakers						
634	2005				CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions				
635	2005		47270	Cardiac resynchronization therapy implantable defibrillator	CT2866	Implantable cardiac electrical stimulators						
636	2001				CT257	Implantable cardiac pacemakers						
637	2003	All other reportable incidents			CT257	Implantable cardiac pacemakers						
638	2003	All other reportable incidents			CT257	Implantable cardiac pacemakers						
639	2003	All other reportable incidents			CT257	Implantable cardiac pacemakers						
640	2003	All other reportable incidents			CT257	Implantable cardiac pacemakers						
641	2001	All other reportable incidents			CT257	Implantable cardiac pacemakers						
642	2001				CT257	Implantable cardiac pacemakers						
643	2001	All other reportable incidents			CT257	Implantable cardiac pacemakers						
644	2005				CT257	Implantable cardiac pacemakers	F1905	Device Revision or Replacement				
645	2001				CT257	Implantable cardiac pacemakers						
646	2003	All other reportable incidents			CT257	Implantable cardiac pacemakers						
647	2005				CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions				
648	2003	All other reportable incidents			CT257	Implantable cardiac pacemakers						
649	2005				CT257	Implantable cardiac pacemakers	F1905	Device Revision or Replacement				
650	2001				CT257	Implantable cardiac pacemakers						
651	2001				CT257	Implantable cardiac pacemakers						
652	2005				CT257	Implantable cardiac pacemakers						
653	2005				CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions				
654	2001	All other reportable incidents			CT257	Implantable cardiac pacemakers						
655	2003	All other reportable incidents			CT257	Implantable cardiac pacemakers						
656	2003	All other reportable incidents			CT257	Implantable cardiac pacemakers						
657	2001	All other reportable incidents			CT257	Implantable cardiac pacemakers						
658	2003	All other reportable incidents			CT257	Implantable cardiac pacemakers						
659	2001				CT257	Implantable cardiac pacemakers						
660	2005				CT257	Implantable cardiac pacemakers	F1905	Device Revision or Replacement				
661	2005				CT257	Implantable cardiac pacemakers	F1905	Device Revision or Replacement				
662	2005				CT257	Implantable cardiac pacemakers	F1905	Device Revision or Replacement				
663	2005				CT257	Implantable cardiac pacemakers	F1905	Device Revision or Replacement				
664	2005				CT257	Implantable cardiac pacemakers	F1905	Device Revision or Replacement				
665	2003	All other reportable incidents			CT257	Implantable cardiac pacemakers						
666	2001				CT257	Implantable cardiac pacemakers						
667	2005				CT257	Implantable cardiac pacemakers	F1905	Device Revision or Replacement				
668	2005				CT257	Implantable cardiac pacemakers	F1905	Device Revision or Replacement				
669	2005				CT257	Implantable cardiac pacemakers	F1905	Device Revision or Replacement				
670	2003	All other reportable incidents			CT257	Implantable cardiac pacemakers						
671	2001	All other reportable incidents			CT257	Implantable cardiac pacemakers						
672	2001				CT257	Implantable cardiac pacemakers						
673	2003	All other reportable incidents			CT257	Implantable cardiac pacemakers						
674	2001	All other reportable incidents			CT257	Implantable cardiac pacemakers						
675	2001	All other reportable incidents			CT257	Implantable cardiac pacemakers						
676	2001				CT257	Implantable cardiac pacemakers						
677	2001				CT257	Implantable cardiac pacemakers						
678	2003	All other reportable incidents			CT257	Implantable cardiac pacemakers						
679	2001				CT257	Implantable cardiac pacemakers						
680	2003	All other reportable incidents			CT257	Implantable cardiac pacemakers						
681	2003				CT257	Implantable cardiac pacemakers						
682	2001				CT257	Implantable cardiac pacemakers						
683	2001	All other reportable incidents			CT257	Implantable cardiac pacemakers						
684	2001				CT257	Implantable cardiac pacemakers						
685	2001				CT257	Implantable cardiac pacemakers						
686	2001				CT257	Implantable cardiac pacemakers						
687	2003	All other reportable incidents			CT257	Implantable cardiac pacemakers						
688	2001	All other reportable incidents			CT257	Implantable cardiac pacemakers						
689	2003	All other reportable incidents			CT257	Implantable cardiac pacemakers						
690	2003	All other reportable incidents			CT257	Implantable cardiac pacemakers						
691	2003	All other reportable incidents			CT257	Implantable cardiac pacemakers						
692	2003	All other reportable incidents			CT257	Implantable cardiac pacemakers	F1905	Device Revision or Replacement				
693	2001	All other reportable incidents			CT257	Implantable cardiac pacemakers						
694	2003	All other reportable incidents			CT257	Implantable cardiac pacemakers						
695	2003	All other reportable incidents			CT257	Implantable cardiac pacemakers						

696	2003	All other reportable incidents			CT257	Implantable cardiac pacemakers						
697	2001				CT257	Implantable cardiac pacemakers						
698	2003	All other reportable incidents			CT257	Implantable cardiac pacemakers						
699	2003	All other reportable incidents			CT257	Implantable cardiac pacemakers						
700	2003	All other reportable incidents			CT257	Implantable cardiac pacemakers						
701	2003	All other reportable incidents			CT257	Implantable cardiac pacemakers						
702	2001				CT257	Implantable cardiac pacemakers						
703	2004	All other reportable incidents			CT257	Implantable cardiac pacemakers						
704	2001	All other reportable incidents			CT257	Implantable cardiac pacemakers						
705	2004	All other reportable incidents			CT257	Implantable cardiac pacemakers						
706	2004	All other reportable incidents			CT257	Implantable cardiac pacemakers						
707	2004	All other reportable incidents			CT257	Implantable cardiac pacemakers	F1905	Device Revision or Replacement				
708	2001				CT257	Implantable cardiac pacemakers						
709	2001	All other reportable incidents			CT257	Implantable cardiac pacemakers						
710	2004	All other reportable incidents			CT257	Implantable cardiac pacemakers	F1905	Device Revision or Replacement				
711	2001	All other reportable incidents			CT257	Implantable cardiac pacemakers						
712	2004	All other reportable incidents			CT257	Implantable cardiac pacemakers						
713	2004	All other reportable incidents			CT257	Implantable cardiac pacemakers						
714	2001				CT257	Implantable cardiac pacemakers						
715	2001				CT257	Implantable cardiac pacemakers						
716	2001	All other reportable incidents			CT257	Implantable cardiac pacemakers						
717	2004	All other reportable incidents			CT257	Implantable cardiac pacemakers						
718	2001				CT257	Implantable cardiac pacemakers						
719	2001				CT257	Implantable cardiac pacemakers						
720	2004	All other reportable incidents			CT257	Implantable cardiac pacemakers						
721	2004	All other reportable incidents			CT257	Implantable cardiac pacemakers	F1905	Device Revision or Replacement				
722	2001				CT257	Implantable cardiac pacemakers						
723	2004	All other reportable incidents			CT257	Implantable cardiac pacemakers						
724	2001				CT257	Implantable cardiac pacemakers						
725	2001				CT257	Implantable cardiac pacemakers						
726	2001	All other reportable incidents			CT257	Implantable cardiac pacemakers						
727	2004	All other reportable incidents			CT257	Implantable cardiac pacemakers						
728	2004	All other reportable incidents			CT257	Implantable cardiac pacemakers						
729	2004	All other reportable incidents			CT257	Implantable cardiac pacemakers						
730	2001	All other reportable incidents			CT257	Implantable cardiac pacemakers						
731	2001				CT257	Implantable cardiac pacemakers						
732	2001	All other reportable incidents			CT257	Implantable cardiac pacemakers						
733	2001	All other reportable incidents			CT257	Implantable cardiac pacemakers						
734	2001				CT257	Implantable cardiac pacemakers						
735	2001				CT257	Implantable cardiac pacemakers						
736	2001				CT257	Implantable cardiac pacemakers						
737	2001				CT257	Implantable cardiac pacemakers						
738	2001	All other reportable incidents			CT257	Implantable cardiac pacemakers						
739	2001				CT257	Implantable cardiac pacemakers						
740	2001				CT257	Implantable cardiac pacemakers						
741	2001				CT257	Implantable cardiac pacemakers						
742	2001				CT257	Implantable cardiac pacemakers						
743	2001				CT257	Implantable cardiac pacemakers						
744	2001	All other reportable incidents			CT257	Implantable cardiac pacemakers						
745	2001	All other reportable incidents			CT257	Implantable cardiac pacemakers						
746	2001				CT257	Implantable cardiac pacemakers						
747	2001				CT257	Implantable cardiac pacemakers						
748	2001				CT257	Implantable cardiac pacemakers						
749	2001				CT257	Implantable cardiac pacemakers						
750	2001	All other reportable incidents			CT257	Implantable cardiac pacemakers						
751	2001				CT257	Implantable cardiac pacemakers						
752	2001	All other reportable incidents			CT257	Implantable cardiac pacemakers						
753	2001	All other reportable incidents			CT257	Implantable cardiac pacemakers						
754	2001	All other reportable incidents			CT257	Implantable cardiac pacemakers						
755	2001	All other reportable incidents			CT257	Implantable cardiac pacemakers						
756	2001	All other reportable incidents			CT257	Implantable cardiac pacemakers						
757	2001	All other reportable incidents			CT257	Implantable cardiac pacemakers						
758	2005		35223	Endocardial/interventricular septal pacing lead	CT1684	Cardiac pacemakers and associated devices	F1905	Device Revision or Replacement	A27	Appropriate Device Problem Term/Code not Available		
759	2008				CT257	Implantable cardiac pacemakers			A27	Appropriate Device Problem Term/Code not Available		
760	2007				CT257	Implantable cardiac pacemakers	F1905	Device Revision or Replacement	A27	Appropriate Device Problem Term/Code not Available		
761	2006				CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A27	Appropriate Device Problem Term/Code not Available		
762	2006		35223	Endocardial/interventricular septal pacing lead	CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A27	Appropriate Device Problem Term/Code not Available		
763	2009	All other reportable incidents			CT257	Implantable cardiac pacemakers			A27	Appropriate Device Problem Term/Code not Available		
764	2008				CT257	Implantable cardiac pacemakers			A27	Appropriate Device Problem Term/Code not Available		
765	2008				CT257	Implantable cardiac pacemakers			A27	Appropriate Device Problem Term/Code not Available		
766	2010	All other reportable incidents			CT257	Implantable cardiac pacemakers			A26	Insufficient Device Problem Information		
767	2010				CT257	Implantable cardiac pacemakers			A27	Appropriate Device Problem Term/Code not Available		
768	2008				CT257	Implantable cardiac pacemakers			A27	Appropriate Device Problem Term/Code not Available		
769	2009				CT257	Implantable cardiac pacemakers			A27	Appropriate Device Problem Term/Code not Available		
770	2008				CT257	Implantable cardiac pacemakers			A27	Appropriate Device Problem Term/Code not Available		
771	2006				CT257	Implantable cardiac pacemakers	F1905	Device Revision or Replacement	A27	Appropriate Device Problem Term/Code not Available		
772	2008				CT257	Implantable cardiac pacemakers			A27	Appropriate Device Problem Term/Code not Available		

773	2007				CT257	Implantable cardiac pacemakers	F1905	Device Revision or Replacement	A27	Appropriate Device Problem Term/Code not Available		
774	2009		35223	Endocardial/interventricular septal pacing lead	CT1684	Cardiac pacemakers and associated devices			A27	Appropriate Device Problem Term/Code not Available		
775	2006				CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A27	Appropriate Device Problem Term/Code not Available		
776	2008				CT257	Implantable cardiac pacemakers			A27	Appropriate Device Problem Term/Code not Available		
777	2009	All other reportable incidents			CT257	Implantable cardiac pacemakers			A27	Appropriate Device Problem Term/Code not Available	12/02/2009	
778	2006				CT257	Implantable cardiac pacemakers	F1905	Device Revision or Replacement	A27	Appropriate Device Problem Term/Code not Available		
779	2006		35223	Endocardial/interventricular septal pacing lead	CT1684	Cardiac pacemakers and associated devices	F1905	Device Revision or Replacement	A27	Appropriate Device Problem Term/Code not Available		
780	2008		47265	Dual-chamber implantable pacemaker, rate-responsive	CT141	Electrodes/Leads and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	05/11/2022	05/11/2022
									A27	Appropriate Device Problem Term/Code not Available	05/11/2022	05/11/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	05/11/2022	05/11/2022
									A27	Appropriate Device Problem Term/Code not Available	05/11/2022	05/11/2022
					CT233	Surgical	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	05/11/2022	05/11/2022
									A27	Appropriate Device Problem Term/Code not Available	05/11/2022	05/11/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	05/11/2022	05/11/2022
									A27	Appropriate Device Problem Term/Code not Available	05/11/2022	05/11/2022
					CT2407	Active implantable	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	05/11/2022	05/11/2022
									A27	Appropriate Device Problem Term/Code not Available	05/11/2022	05/11/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	05/11/2022	05/11/2022
									A27	Appropriate Device Problem Term/Code not Available	05/11/2022	05/11/2022
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	05/11/2022	05/11/2022
									A27	Appropriate Device Problem Term/Code not Available	05/11/2022	05/11/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	05/11/2022	05/11/2022
									A27	Appropriate Device Problem Term/Code not Available	05/11/2022	05/11/2022
					CT3001	Implantable metal	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	05/11/2022	05/11/2022
									A27	Appropriate Device Problem Term/Code not Available	05/11/2022	05/11/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	05/11/2022	05/11/2022
									A27	Appropriate Device Problem Term/Code not Available	05/11/2022	05/11/2022
					CT3002	Implantable synthetic polymer	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	05/11/2022	05/11/2022
									A27	Appropriate Device Problem Term/Code not Available	05/11/2022	05/11/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	05/11/2022	05/11/2022
									A27	Appropriate Device Problem Term/Code not Available	05/11/2022	05/11/2022
					CT314	Electrically powered	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	05/11/2022	05/11/2022
									A27	Appropriate Device Problem Term/Code not Available	05/11/2022	05/11/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	05/11/2022	05/11/2022
									A27	Appropriate Device Problem Term/Code not Available	05/11/2022	05/11/2022
					CT321	Long-term surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	05/11/2022	05/11/2022
									A27	Appropriate Device Problem Term/Code not Available	05/11/2022	05/11/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	05/11/2022	05/11/2022
									A27	Appropriate Device Problem Term/Code not Available	05/11/2022	05/11/2022
					CT449	Generators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	05/11/2022	05/11/2022
									A27	Appropriate Device Problem Term/Code not Available	05/11/2022	05/11/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	05/11/2022	05/11/2022
									A27	Appropriate Device Problem Term/Code not Available	05/11/2022	05/11/2022
					CT981	Single use	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	05/11/2022	05/11/2022
									A27	Appropriate Device Problem Term/Code not Available	05/11/2022	05/11/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	05/11/2022	05/11/2022
									A27	Appropriate Device Problem Term/Code not Available	05/11/2022	05/11/2022
781	2006				CT257	Implantable cardiac pacemakers	F1905	Device Revision or Replacement	A27	Appropriate Device Problem Term/Code not Available		
782	2009				CT257	Implantable cardiac pacemakers			A27	Appropriate Device Problem Term/Code not Available		
783	2006				CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A27	Appropriate Device Problem Term/Code not Available		
784	2006				CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A27	Appropriate Device Problem Term/Code not Available		
785	2010	All other reportable incidents			CT257	Implantable cardiac pacemakers			A27	Appropriate Device Problem Term/Code not Available		
786	2008		35223	Endocardial/interventricular septal pacing lead	CT1684	Cardiac pacemakers and associated devices			A27	Appropriate Device Problem Term/Code not Available		
787	2007				CT257	Implantable cardiac pacemakers	F1905	Device Revision or Replacement	A27	Appropriate Device Problem Term/Code not Available		
788	2008		47270	Cardiac resynchronization therapy implantable defibrillator	CT2866	Implantable cardiac electrical stimulators			A26	Insufficient Device Problem Information		
789	2010		47270	Cardiac resynchronization therapy implantable defibrillator	CT2866	Implantable cardiac electrical stimulators			A27	Appropriate Device Problem Term/Code not Available		
790	2006				CT257	Implantable cardiac pacemakers	F1905	Device Revision or Replacement	A27	Appropriate Device Problem Term/Code not Available		
791	2010		35223	Endocardial/interventricular septal pacing lead	CT1684	Cardiac pacemakers and associated devices			A26	Insufficient Device Problem Information		

792	2008		47270	Cardiac resynchronization therapy implantable defibrillator	CT2866	Implantable cardiac electrical stimulators			A27	Appropriate Device Problem Term/Code not Available		
793	2008		47270	Cardiac resynchronization therapy implantable defibrillator	CT2866	Implantable cardiac electrical stimulators			A27	Appropriate Device Problem Term/Code not Available		
794	2008		47270	Cardiac resynchronization therapy implantable defibrillator	CT2866	Implantable cardiac electrical stimulators			A27	Appropriate Device Problem Term/Code not Available		
795	2006		35223	Endocardial/interventricular septal pacing lead	CT1684	Cardiac pacemakers and associated devices	F1905	Device Revision or Replacement	A27	Appropriate Device Problem Term/Code not Available		
796	2006		35223	Endocardial/interventricular septal pacing lead	CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A27	Appropriate Device Problem Term/Code not Available		
797	2010				CT257	Implantable cardiac pacemakers			A27	Appropriate Device Problem Term/Code not Available		
798	2009				CT257	Implantable cardiac pacemakers			A27	Appropriate Device Problem Term/Code not Available		
799	2008		47270	Cardiac resynchronization therapy implantable defibrillator	CT2866	Implantable cardiac electrical stimulators			A27	Appropriate Device Problem Term/Code not Available		
800	2009		47270	Cardiac resynchronization therapy implantable defibrillator	CT2866	Implantable cardiac electrical stimulators			A26	Insufficient Device Problem Information		
801	2006				CT257	Implantable cardiac pacemakers	F1905	Device Revision or Replacement	A27	Appropriate Device Problem Term/Code not Available		
802	2010				CT257	Implantable cardiac pacemakers			A27	Appropriate Device Problem Term/Code not Available		
803	2006				CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A27	Appropriate Device Problem Term/Code not Available		
804	2009				CT257	Implantable cardiac pacemakers			A27	Appropriate Device Problem Term/Code not Available		
805	2009				CT257	Implantable cardiac pacemakers			A27	Appropriate Device Problem Term/Code not Available		
806	2008				CT257	Implantable cardiac pacemakers			A27	Appropriate Device Problem Term/Code not Available		
807	2010		47270	Cardiac resynchronization therapy implantable defibrillator	CT2866	Implantable cardiac electrical stimulators			A27	Appropriate Device Problem Term/Code not Available		
808	2009				CT257	Implantable cardiac pacemakers			A27	Appropriate Device Problem Term/Code not Available		
809	2006				CT257	Implantable cardiac pacemakers			A27	Appropriate Device Problem Term/Code not Available		
810	2008				CT257	Implantable cardiac pacemakers			A27	Appropriate Device Problem Term/Code not Available		
811	2008				CT257	Implantable cardiac pacemakers			A27	Appropriate Device Problem Term/Code not Available		
812	2010				CT257	Implantable cardiac pacemakers			A27	Appropriate Device Problem Term/Code not Available		
813	2008				CT257	Implantable cardiac pacemakers			A26	Insufficient Device Problem Information		
814	2008				CT257	Implantable cardiac pacemakers			A27	Appropriate Device Problem Term/Code not Available		
815	2008		35223	Endocardial/interventricular septal pacing lead	CT1684	Cardiac pacemakers and associated devices			A27	Appropriate Device Problem Term/Code not Available		
816	2009	All other reportable incidents			CT257	Implantable cardiac pacemakers			A27	Appropriate Device Problem Term/Code not Available		
817	2009	All other reportable incidents			CT257	Implantable cardiac pacemakers			A27	Appropriate Device Problem Term/Code not Available		
818	2009				CT257	Implantable cardiac pacemakers			A27	Appropriate Device Problem Term/Code not Available		
819	2010		35223	Endocardial/interventricular septal pacing lead	CT1684	Cardiac pacemakers and associated devices			A27	Appropriate Device Problem Term/Code not Available		
820	2008		47270	Cardiac resynchronization therapy implantable defibrillator	CT2866	Implantable cardiac electrical stimulators			A26	Insufficient Device Problem Information		
821	2007				CT257	Implantable cardiac pacemakers			A26	Insufficient Device Problem Information		
822	2008				CT257	Implantable cardiac pacemakers			A26	Insufficient Device Problem Information		
823	2007				CT257	Implantable cardiac pacemakers	F1905	Device Revision or Replacement	A27	Appropriate Device Problem Term/Code not Available		
824	2009				CT257	Implantable cardiac pacemakers			A27	Appropriate Device Problem Term/Code not Available		
825	2009				CT257	Implantable cardiac pacemakers			A27	Appropriate Device Problem Term/Code not Available		
826	2010				CT257	Implantable cardiac pacemakers			A27	Appropriate Device Problem Term/Code not Available		
827	2010				CT257	Implantable cardiac pacemakers			A27	Appropriate Device Problem Term/Code not Available		
828	2009		35223	Endocardial/interventricular septal pacing lead	CT1684	Cardiac pacemakers and associated devices			A27	Appropriate Device Problem Term/Code not Available		
829	2010				CT257	Implantable cardiac pacemakers			A27	Appropriate Device Problem Term/Code not Available		
830	2008				CT257	Implantable cardiac pacemakers			A27	Appropriate Device Problem Term/Code not Available		
831	2010				CT257	Implantable cardiac pacemakers			A27	Appropriate Device Problem Term/Code not Available		
832	2009				CT257	Implantable cardiac pacemakers			A26	Insufficient Device Problem Information		
833	2010		47270	Cardiac resynchronization therapy implantable defibrillator	CT2866	Implantable cardiac electrical stimulators			A26	Insufficient Device Problem Information		
834	2009		47270	Cardiac resynchronization therapy implantable defibrillator	CT2866	Implantable cardiac electrical stimulators			A26	Insufficient Device Problem Information		
835	2008	All other reportable incidents			CT257	Implantable cardiac pacemakers			A27	Appropriate Device Problem Term/Code not Available		
836	2006				CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A27	Appropriate Device Problem Term/Code not Available		
837	2010				CT257	Implantable cardiac pacemakers			A27	Appropriate Device Problem Term/Code not Available		
838	2007		35223	Endocardial/interventricular septal pacing lead	CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A26	Insufficient Device Problem Information		
839	2008				CT257	Implantable cardiac pacemakers			A27	Appropriate Device Problem Term/Code not Available		
840	2008				CT257	Implantable cardiac pacemakers			A27	Appropriate Device Problem Term/Code not Available		
841	2006		60910	Coronary venous pacing lead	CT1684	Cardiac pacemakers and associated devices	F1905	Device Revision or Replacement	A27	Appropriate Device Problem Term/Code not Available		
842	2009				CT257	Implantable cardiac pacemakers			A27	Appropriate Device Problem Term/Code not Available		
843	2007		35039	Epicardial pacing lead	CT1684	Cardiac pacemakers and associated devices			A26	Insufficient Device Problem Information		

844	2007				CT257	Implantable cardiac pacemakers			A27	Appropriate Device Problem Term/Code not Available		
845	2007				CT257	Implantable cardiac pacemakers	F1905	Device Revision or Replacement	A27	Appropriate Device Problem Term/Code not Available		
846	2006				CT257	Implantable cardiac pacemakers	F1905	Device Revision or Replacement	A27	Appropriate Device Problem Term/Code not Available		
847	2009		47270	Cardiac resynchronization therapy implantable defibrillator	CT2866	Implantable cardiac electrical stimulators			A27	Appropriate Device Problem Term/Code not Available		
848	2007				CT257	Implantable cardiac pacemakers	F1905	Device Revision or Replacement	A27	Appropriate Device Problem Term/Code not Available		
849	2010		47270	Cardiac resynchronization therapy implantable defibrillator	CT2866	Implantable cardiac electrical stimulators			A27	Appropriate Device Problem Term/Code not Available		
850	2006		47270	Cardiac resynchronization therapy implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	F1905	Device Revision or Replacement	A26	Insufficient Device Problem Information		
851	2009				CT257	Implantable cardiac pacemakers			A27	Appropriate Device Problem Term/Code not Available		
852	2010				CT257	Implantable cardiac pacemakers			A27	Appropriate Device Problem Term/Code not Available		
853	2010		47270	Cardiac resynchronization therapy implantable defibrillator	CT2866	Implantable cardiac electrical stimulators			A27	Appropriate Device Problem Term/Code not Available		
854	2009	All other reportable incidents			CT257	Implantable cardiac pacemakers			A27	Appropriate Device Problem Term/Code not Available		
855	2010		47270	Cardiac resynchronization therapy implantable defibrillator	CT2866	Implantable cardiac electrical stimulators			A27	Appropriate Device Problem Term/Code not Available		
856	2009				CT257	Implantable cardiac pacemakers			A27	Appropriate Device Problem Term/Code not Available		
857	2009		35223	Endocardial/interventricular septal pacing lead	CT1684	Cardiac pacemakers and associated devices			A27	Appropriate Device Problem Term/Code not Available		
858	2007				CT257	Implantable cardiac pacemakers			A27	Appropriate Device Problem Term/Code not Available		
859	2010		47270	Cardiac resynchronization therapy implantable defibrillator	CT2866	Implantable cardiac electrical stimulators			A27	Appropriate Device Problem Term/Code not Available		
860	2009		47270	Cardiac resynchronization therapy implantable defibrillator	CT2866	Implantable cardiac electrical stimulators			A27	Appropriate Device Problem Term/Code not Available		
861	2009				CT257	Implantable cardiac pacemakers			A27	Appropriate Device Problem Term/Code not Available		
862	2006				CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A27	Appropriate Device Problem Term/Code not Available		
863	2009				CT257	Implantable cardiac pacemakers			A27	Appropriate Device Problem Term/Code not Available		
864	2009				CT257	Implantable cardiac pacemakers			A27	Appropriate Device Problem Term/Code not Available		
865	2007		35223	Endocardial/interventricular septal pacing lead	CT1684	Cardiac pacemakers and associated devices	F1905	Device Revision or Replacement	A27	Appropriate Device Problem Term/Code not Available		
866	2006				CT257	Implantable cardiac pacemakers	F1905	Device Revision or Replacement	A27	Appropriate Device Problem Term/Code not Available		
867	2007				CT257	Implantable cardiac pacemakers			A27	Appropriate Device Problem Term/Code not Available		
868	2010				CT257	Implantable cardiac pacemakers			A27	Appropriate Device Problem Term/Code not Available		
869	2009	All other reportable incidents			CT257	Implantable cardiac pacemakers			A27	Appropriate Device Problem Term/Code not Available		
870	2010				CT257	Implantable cardiac pacemakers			A27	Appropriate Device Problem Term/Code not Available		
871	2009				CT257	Implantable cardiac pacemakers			A27	Appropriate Device Problem Term/Code not Available		
872	2010		47270	Cardiac resynchronization therapy implantable defibrillator	CT2866	Implantable cardiac electrical stimulators			A27	Appropriate Device Problem Term/Code not Available		
873	2009				CT257	Implantable cardiac pacemakers			A27	Appropriate Device Problem Term/Code not Available		
874	2010		47270	Cardiac resynchronization therapy implantable defibrillator	CT2866	Implantable cardiac electrical stimulators			A27	Appropriate Device Problem Term/Code not Available		
875	2006				CT257	Implantable cardiac pacemakers	F1905	Device Revision or Replacement	A27	Appropriate Device Problem Term/Code not Available		
876	2010		47270	Cardiac resynchronization therapy implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	F1905	Device Revision or Replacement	A27	Appropriate Device Problem Term/Code not Available		
877	2006				CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A27	Appropriate Device Problem Term/Code not Available		
878	2007		47270	Cardiac resynchronization therapy implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	F1905	Device Revision or Replacement	A27	Appropriate Device Problem Term/Code not Available		
879	2007				CT257	Implantable cardiac pacemakers			A27	Appropriate Device Problem Term/Code not Available		
880	2006		47270	Cardiac resynchronization therapy implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	F1905	Device Revision or Replacement	A27	Appropriate Device Problem Term/Code not Available		
881	2009	All other reportable incidents			CT257	Implantable cardiac pacemakers			A27	Appropriate Device Problem Term/Code not Available		
882	2009				CT257	Implantable cardiac pacemakers			A27	Appropriate Device Problem Term/Code not Available		
883	2008				CT257	Implantable cardiac pacemakers			A27	Appropriate Device Problem Term/Code not Available		
884	2010				CT257	Implantable cardiac pacemakers			A27	Appropriate Device Problem Term/Code not Available		
885	2009				CT257	Implantable cardiac pacemakers			A27	Appropriate Device Problem Term/Code not Available		
886	2009				CT257	Implantable cardiac pacemakers			A27	Appropriate Device Problem Term/Code not Available		
887	2009				CT257	Implantable cardiac pacemakers			A26	Insufficient Device Problem Information		
888	2007				CT257	Implantable cardiac pacemakers			A27	Appropriate Device Problem Term/Code not Available		
889	2006				CT257	Implantable cardiac pacemakers	F1905	Device Revision or Replacement	A27	Appropriate Device Problem Term/Code not Available		
890	2006		35223	Endocardial/interventricular septal pacing lead	CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A27	Appropriate Device Problem Term/Code not Available		
891	2008				CT257	Implantable cardiac pacemakers			A27	Appropriate Device Problem Term/Code not Available		
892	2010		47270	Cardiac resynchronization therapy implantable defibrillator	CT2866	Implantable cardiac electrical stimulators			A26	Insufficient Device Problem Information		
893	2008				CT257	Implantable cardiac pacemakers			A27	Appropriate Device Problem Term/Code not Available		

894	2010		47270	Cardiac resynchronization therapy implantable defibrillator	CT2866	Implantable cardiac electrical stimulators			A27	Appropriate Device Problem Term/Code not Available		
895	2009				CT257	Implantable cardiac pacemakers			A27	Appropriate Device Problem Term/Code not Available		
896	2007		35223	Endocardial/interventricular septal pacing lead	CT1684	Cardiac pacemakers and associated devices	F1905	Device Revision or Replacement	A27	Appropriate Device Problem Term/Code not Available		
897	2006		47270	Cardiac resynchronization therapy implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	F1905	Device Revision or Replacement	A27	Appropriate Device Problem Term/Code not Available		
898	2009				CT257	Implantable cardiac pacemakers			A26	Insufficient Device Problem Information		
899	2009		47270	Cardiac resynchronization therapy implantable defibrillator	CT2866	Implantable cardiac electrical stimulators			A26	Insufficient Device Problem Information		
900	2006				CT257	Implantable cardiac pacemakers	F1905	Device Revision or Replacement	A27	Appropriate Device Problem Term/Code not Available		
901	2007		35223	Endocardial/interventricular septal pacing lead	CT1684	Cardiac pacemakers and associated devices	F1905	Device Revision or Replacement	A27	Appropriate Device Problem Term/Code not Available		
902	2007		35223	Endocardial/interventricular septal pacing lead	CT1684	Cardiac pacemakers and associated devices			A27	Appropriate Device Problem Term/Code not Available		
903	2009				CT257	Implantable cardiac pacemakers			A27	Appropriate Device Problem Term/Code not Available		
904	2009				CT257	Implantable cardiac pacemakers			A27	Appropriate Device Problem Term/Code not Available		
905	2010				CT257	Implantable cardiac pacemakers			A26	Insufficient Device Problem Information		
906	2008				CT257	Implantable cardiac pacemakers			A27	Appropriate Device Problem Term/Code not Available		
907	2009		47270	Cardiac resynchronization therapy implantable defibrillator	CT2866	Implantable cardiac electrical stimulators			A27	Appropriate Device Problem Term/Code not Available		
908	2006				CT257	Implantable cardiac pacemakers	F1905	Device Revision or Replacement	A27	Appropriate Device Problem Term/Code not Available		
909	2009		35223	Endocardial/interventricular septal pacing lead	CT1684	Cardiac pacemakers and associated devices			A27	Appropriate Device Problem Term/Code not Available		
910	2008		47270	Cardiac resynchronization therapy implantable defibrillator	CT2866	Implantable cardiac electrical stimulators			A27	Appropriate Device Problem Term/Code not Available		
911	2006				CT257	Implantable cardiac pacemakers	F1905	Device Revision or Replacement	A27	Appropriate Device Problem Term/Code not Available		
912	2007	All other reportable incidents			CT257	Implantable cardiac pacemakers			A27	Appropriate Device Problem Term/Code not Available		
913	2006				CT257	Implantable cardiac pacemakers	F1905	Device Revision or Replacement	A27	Appropriate Device Problem Term/Code not Available		
914	2006				CT257	Implantable cardiac pacemakers	F1905	Device Revision or Replacement	A27	Appropriate Device Problem Term/Code not Available		
915	2008				CT257	Implantable cardiac pacemakers			A26	Insufficient Device Problem Information		
916	2010				CT257	Implantable cardiac pacemakers			A27	Appropriate Device Problem Term/Code not Available		
917	2006		60910	Coronary venous pacing lead	CT1684	Cardiac pacemakers and associated devices	F1905	Device Revision or Replacement	A27	Appropriate Device Problem Term/Code not Available		
918	2010				CT257	Implantable cardiac pacemakers	E1906	Unspecified Infection	A26	Insufficient Device Problem Information		
919	2008				CT257	Implantable cardiac pacemakers			A27	Appropriate Device Problem Term/Code not Available		
920	2010				CT257	Implantable cardiac pacemakers			A27	Appropriate Device Problem Term/Code not Available		
921	2007				CT257	Implantable cardiac pacemakers	F1905	Device Revision or Replacement	A27	Appropriate Device Problem Term/Code not Available		
922	2008		36104	Implantable pacemaker analyser	CT1684	Cardiac pacemakers and associated devices			A27	Appropriate Device Problem Term/Code not Available		
923	2008				CT257	Implantable cardiac pacemakers			A27	Appropriate Device Problem Term/Code not Available		
924	2009		47270	Cardiac resynchronization therapy implantable defibrillator	CT2866	Implantable cardiac electrical stimulators			A27	Appropriate Device Problem Term/Code not Available		
925	2010				CT257	Implantable cardiac pacemakers	F1905	Device Revision or Replacement	A26	Insufficient Device Problem Information		
926	2009		47270	Cardiac resynchronization therapy implantable defibrillator	CT2866	Implantable cardiac electrical stimulators			A27	Appropriate Device Problem Term/Code not Available		
927	2008				CT257	Implantable cardiac pacemakers			A27	Appropriate Device Problem Term/Code not Available		
928	2007		47270	Cardiac resynchronization therapy implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	F1905	Device Revision or Replacement	A27	Appropriate Device Problem Term/Code not Available		
929	2006				CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A27	Appropriate Device Problem Term/Code not Available		
930	2010		47270	Cardiac resynchronization therapy implantable defibrillator	CT2866	Implantable cardiac electrical stimulators			A27	Appropriate Device Problem Term/Code not Available		
931	2010		35223	Endocardial/interventricular septal pacing lead	CT1684	Cardiac pacemakers and associated devices			A26	Insufficient Device Problem Information		
932	2006				CT257	Implantable cardiac pacemakers			A27	Appropriate Device Problem Term/Code not Available		
933	2009		35223	Endocardial/interventricular septal pacing lead	CT1684	Cardiac pacemakers and associated devices			A27	Appropriate Device Problem Term/Code not Available		
934	2006				CT257	Implantable cardiac pacemakers	F1905	Device Revision or Replacement	A27	Appropriate Device Problem Term/Code not Available		
935	2007				CT257	Implantable cardiac pacemakers	F1905	Device Revision or Replacement	A27	Appropriate Device Problem Term/Code not Available		
936	2006				CT257	Implantable cardiac pacemakers	F1905	Device Revision or Replacement	A27	Appropriate Device Problem Term/Code not Available		
937	2010				CT257	Implantable cardiac pacemakers			A26	Insufficient Device Problem Information		
938	2008				CT257	Implantable cardiac pacemakers			A26	Insufficient Device Problem Information		
939	2010				CT257	Implantable cardiac pacemakers			A27	Appropriate Device Problem Term/Code not Available		
940	2009				CT257	Implantable cardiac pacemakers			A27	Appropriate Device Problem Term/Code not Available		
941	2010		47270	Cardiac resynchronization therapy implantable defibrillator	CT2866	Implantable cardiac electrical stimulators			A27	Appropriate Device Problem Term/Code not Available		
942	2010				CT257	Implantable cardiac pacemakers			A27	Appropriate Device Problem Term/Code not Available		
943	2008				CT257	Implantable cardiac pacemakers			A27	Appropriate Device Problem Term/Code not Available		
944	2006				CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A27	Appropriate Device Problem Term/Code not Available		
945	2010				CT257	Implantable cardiac pacemakers			A27	Appropriate Device Problem Term/Code not Available		
946	2006				CT257	Implantable cardiac pacemakers	F1905	Device Revision or Replacement	A27	Appropriate Device Problem Term/Code not Available		

947	2009				CT257	Implantable cardiac pacemakers			A27	Appropriate Device Problem Term/Code not Available		
948	2006				CT257	Implantable cardiac pacemakers	F1905	Device Revision or Replacement	A27	Appropriate Device Problem Term/Code not Available		
949	2010		35223	Endocardial/interventricular septal pacing lead	CT1684	Cardiac pacemakers and associated devices			A26	Insufficient Device Problem Information		
950	2009				CT257	Implantable cardiac pacemakers			A27	Appropriate Device Problem Term/Code not Available		
951	2009		35223	Endocardial/interventricular septal pacing lead	CT1684	Cardiac pacemakers and associated devices			A27	Appropriate Device Problem Term/Code not Available		
952	2007		35223	Endocardial/interventricular septal pacing lead	CT1684	Cardiac pacemakers and associated devices	F1905	Device Revision or Replacement	A27	Appropriate Device Problem Term/Code not Available		
953	2010		47270	Cardiac resynchronization therapy implantable defibrillator	CT2866	Implantable cardiac electrical stimulators			A27	Appropriate Device Problem Term/Code not Available		
954	2008				CT257	Implantable cardiac pacemakers			A27	Appropriate Device Problem Term/Code not Available		
955	2009				CT257	Implantable cardiac pacemakers			A27	Appropriate Device Problem Term/Code not Available		
956	2010				CT257	Implantable cardiac pacemakers			A26	Insufficient Device Problem Information		
957	2010		47270	Cardiac resynchronization therapy implantable defibrillator	CT2866	Implantable cardiac electrical stimulators			A27	Appropriate Device Problem Term/Code not Available		
958	2010		47270	Cardiac resynchronization therapy implantable defibrillator	CT2866	Implantable cardiac electrical stimulators			A27	Appropriate Device Problem Term/Code not Available		
959	2006				CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A27	Appropriate Device Problem Term/Code not Available		
960	2010				CT257	Implantable cardiac pacemakers			A26	Insufficient Device Problem Information		
961	2010				CT257	Implantable cardiac pacemakers			A27	Appropriate Device Problem Term/Code not Available		
962	2008				CT257	Implantable cardiac pacemakers			A27	Appropriate Device Problem Term/Code not Available		
963	2008		47270	Cardiac resynchronization therapy implantable defibrillator	CT2866	Implantable cardiac electrical stimulators			A27	Appropriate Device Problem Term/Code not Available		
964	2009				CT257	Implantable cardiac pacemakers			A27	Appropriate Device Problem Term/Code not Available		
965	2009	All other reportable incidents			CT257	Implantable cardiac pacemakers			A26	Insufficient Device Problem Information		
966	2008		35223	Endocardial/interventricular septal pacing lead	CT1684	Cardiac pacemakers and associated devices			A27	Appropriate Device Problem Term/Code not Available		
967	2009				CT257	Implantable cardiac pacemakers			A27	Appropriate Device Problem Term/Code not Available		
968	2009				CT257	Implantable cardiac pacemakers			A27	Appropriate Device Problem Term/Code not Available		
969	2007				CT257	Implantable cardiac pacemakers	F1905	Device Revision or Replacement	A27	Appropriate Device Problem Term/Code not Available		
970	2010		35223	Endocardial/interventricular septal pacing lead	CT1684	Cardiac pacemakers and associated devices			A27	Appropriate Device Problem Term/Code not Available		
971	2009				CT257	Implantable cardiac pacemakers			A27	Appropriate Device Problem Term/Code not Available		
972	2008	All other reportable incidents			CT257	Implantable cardiac pacemakers			A27	Appropriate Device Problem Term/Code not Available		
973	2009		35223	Endocardial/interventricular septal pacing lead	CT1684	Cardiac pacemakers and associated devices			A27	Appropriate Device Problem Term/Code not Available		
974	2009				CT257	Implantable cardiac pacemakers			A27	Appropriate Device Problem Term/Code not Available		
975	2009				CT257	Implantable cardiac pacemakers			A27	Appropriate Device Problem Term/Code not Available		
976	2006		35223	Endocardial/interventricular septal pacing lead	CT1684	Cardiac pacemakers and associated devices			A27	Appropriate Device Problem Term/Code not Available		
977	2010				CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A27	Appropriate Device Problem Term/Code not Available		
978	2008		35223	Endocardial/interventricular septal pacing lead	CT1684	Cardiac pacemakers and associated devices			A27	Appropriate Device Problem Term/Code not Available		
979	2008				CT257	Implantable cardiac pacemakers			A27	Appropriate Device Problem Term/Code not Available		
980	2008				CT257	Implantable cardiac pacemakers			A27	Appropriate Device Problem Term/Code not Available		
981	2008				CT257	Implantable cardiac pacemakers			A27	Appropriate Device Problem Term/Code not Available		
982	2006	All other reportable incidents			CT257	Implantable cardiac pacemakers			A27	Appropriate Device Problem Term/Code not Available		
983	2010				CT257	Implantable cardiac pacemakers			A26	Insufficient Device Problem Information		
984	2007				CT257	Implantable cardiac pacemakers	F1905	Device Revision or Replacement	A27	Appropriate Device Problem Term/Code not Available		
985	2006				CT257	Implantable cardiac pacemakers	F1905	Device Revision or Replacement	A27	Appropriate Device Problem Term/Code not Available		
986	2007				CT257	Implantable cardiac pacemakers			A27	Appropriate Device Problem Term/Code not Available		
987	2010				CT257	Implantable cardiac pacemakers			A26	Insufficient Device Problem Information		
988	2006				CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A27	Appropriate Device Problem Term/Code not Available		
989	2008				CT257	Implantable cardiac pacemakers			A27	Appropriate Device Problem Term/Code not Available		
990	2007		47270	Cardiac resynchronization therapy implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	F1905	Device Revision or Replacement	A27	Appropriate Device Problem Term/Code not Available		
991	2010				CT257	Implantable cardiac pacemakers			A27	Appropriate Device Problem Term/Code not Available		
992	2009		47270	Cardiac resynchronization therapy implantable defibrillator	CT2866	Implantable cardiac electrical stimulators			A27	Appropriate Device Problem Term/Code not Available		
993	2006	All other reportable incidents			CT257	Implantable cardiac pacemakers			A27	Appropriate Device Problem Term/Code not Available		
994	2008				CT257	Implantable cardiac pacemakers			A27	Appropriate Device Problem Term/Code not Available		
995	2010				CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A27	Appropriate Device Problem Term/Code not Available		
996	2008				CT257	Implantable cardiac pacemakers			A27	Appropriate Device Problem Term/Code not Available		
997	2007				CT257	Implantable cardiac pacemakers	F1905	Device Revision or Replacement	A27	Appropriate Device Problem Term/Code not Available		
998	2007				CT257	Implantable cardiac pacemakers	F1905	Device Revision or Replacement	A27	Appropriate Device Problem Term/Code not Available		

999	2009	Unanticipated serious deterioration in state of health			CT257	Implantable cardiac pacemakers			A26	Insufficient Device Problem Information		
1000	2008				CT257	Implantable cardiac pacemakers			A27	Appropriate Device Problem Term/Code not Available		
1001	2007		35223	Endocardial/interventricular septal pacing lead	CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A26	Insufficient Device Problem Information		
1002	2008				CT257	Implantable cardiac pacemakers			A27	Appropriate Device Problem Term/Code not Available		
1003	2007				CT257	Implantable cardiac pacemakers	F1905	Device Revision or Replacement	A27	Appropriate Device Problem Term/Code not Available		
1004	2009				CT257	Implantable cardiac pacemakers			A27	Appropriate Device Problem Term/Code not Available		
1005	2010				CT257	Implantable cardiac pacemakers			A27	Appropriate Device Problem Term/Code not Available		
1006	2010		35223	Endocardial/interventricular septal pacing lead	CT1684	Cardiac pacemakers and associated devices			A27	Appropriate Device Problem Term/Code not Available		
1007	2006				CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A27	Appropriate Device Problem Term/Code not Available		
1008	2010		47270	Cardiac resynchronization therapy implantable defibrillator	CT2866	Implantable cardiac electrical stimulators			A27	Appropriate Device Problem Term/Code not Available		
1009	2008				CT257	Implantable cardiac pacemakers			A27	Appropriate Device Problem Term/Code not Available		
1010	2007		47270	Cardiac resynchronization therapy implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	F1905	Device Revision or Replacement	A27	Appropriate Device Problem Term/Code not Available		
1011	2007		47270	Cardiac resynchronization therapy implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	F1905	Device Revision or Replacement	A27	Appropriate Device Problem Term/Code not Available		
1012	2006				CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A27	Appropriate Device Problem Term/Code not Available		
1013	2009				CT257	Implantable cardiac pacemakers			A27	Appropriate Device Problem Term/Code not Available		
1014	2009				CT257	Implantable cardiac pacemakers			A27	Appropriate Device Problem Term/Code not Available		
1015	2008				CT257	Implantable cardiac pacemakers			A27	Appropriate Device Problem Term/Code not Available		
1016	2007		35223	Endocardial/interventricular septal pacing lead	CT1684	Cardiac pacemakers and associated devices			A27	Appropriate Device Problem Term/Code not Available		
1017	2010		47270	Cardiac resynchronization therapy implantable defibrillator	CT2866	Implantable cardiac electrical stimulators			A27	Appropriate Device Problem Term/Code not Available		
1018	2008				CT257	Implantable cardiac pacemakers			A27	Appropriate Device Problem Term/Code not Available		
1019	2006				CT257	Implantable cardiac pacemakers			A27	Appropriate Device Problem Term/Code not Available		
1020	2006	All other reportable incidents			CT257	Implantable cardiac pacemakers			A27	Appropriate Device Problem Term/Code not Available		
1021	2008	All other reportable incidents			CT257	Implantable cardiac pacemakers			A27	Appropriate Device Problem Term/Code not Available		
1022	2009				CT257	Implantable cardiac pacemakers			A27	Appropriate Device Problem Term/Code not Available		
1023	2009				CT257	Implantable cardiac pacemakers			A27	Appropriate Device Problem Term/Code not Available		
1024	2009		47270	Cardiac resynchronization therapy implantable defibrillator	CT2866	Implantable cardiac electrical stimulators			A27	Appropriate Device Problem Term/Code not Available		
1025	2007				CT257	Implantable cardiac pacemakers	F1905	Device Revision or Replacement	A26	Insufficient Device Problem Information		
1026	2007				CT257	Implantable cardiac pacemakers	F1905	Device Revision or Replacement	A27	Appropriate Device Problem Term/Code not Available		
1027	2009	All other reportable incidents			CT257	Implantable cardiac pacemakers			A26	Insufficient Device Problem Information		
1028	2008				CT257	Implantable cardiac pacemakers			A27	Appropriate Device Problem Term/Code not Available		
1029	2006				CT257	Implantable cardiac pacemakers	F1905	Device Revision or Replacement	A27	Appropriate Device Problem Term/Code not Available		
1030	2006		35223	Endocardial/interventricular septal pacing lead	CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A27	Appropriate Device Problem Term/Code not Available		
1031	2006				CT257	Implantable cardiac pacemakers	F1905	Device Revision or Replacement	A27	Appropriate Device Problem Term/Code not Available		
1032	2006				CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A27	Appropriate Device Problem Term/Code not Available		
1033	2010				CT257	Implantable cardiac pacemakers			A27	Appropriate Device Problem Term/Code not Available		
1034	2008				CT257	Implantable cardiac pacemakers			A27	Appropriate Device Problem Term/Code not Available		
1035	2008		47270	Cardiac resynchronization therapy implantable defibrillator	CT2866	Implantable cardiac electrical stimulators			A27	Appropriate Device Problem Term/Code not Available		
1036	2009		47270	Cardiac resynchronization therapy implantable defibrillator	CT2866	Implantable cardiac electrical stimulators			A27	Appropriate Device Problem Term/Code not Available		
1037	2008				CT257	Implantable cardiac pacemakers			A26	Insufficient Device Problem Information		
1038	2009				CT257	Implantable cardiac pacemakers			A27	Appropriate Device Problem Term/Code not Available		
1039	2008				CT257	Implantable cardiac pacemakers			A27	Appropriate Device Problem Term/Code not Available		
1040	2006	All other reportable incidents			CT257	Implantable cardiac pacemakers	F1905	Device Revision or Replacement	A27	Appropriate Device Problem Term/Code not Available		
1041	2008		47270	Cardiac resynchronization therapy implantable defibrillator	CT2866	Implantable cardiac electrical stimulators			A27	Appropriate Device Problem Term/Code not Available		
1042	2010		60910	Coronary venous pacing lead	CT1684	Cardiac pacemakers and associated devices			A26	Insufficient Device Problem Information		
1043	2006				CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A27	Appropriate Device Problem Term/Code not Available		
1044	2006				CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A27	Appropriate Device Problem Term/Code not Available		
1045	2009		47270	Cardiac resynchronization therapy implantable defibrillator	CT2866	Implantable cardiac electrical stimulators			A27	Appropriate Device Problem Term/Code not Available		
1046	2007				CT257	Implantable cardiac pacemakers	F1905	Device Revision or Replacement	A27	Appropriate Device Problem Term/Code not Available		
1047	2010				CT257	Implantable cardiac pacemakers			A27	Appropriate Device Problem Term/Code not Available		
1048	2007		47270	Cardiac resynchronization therapy implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	F1905	Device Revision or Replacement	A27	Appropriate Device Problem Term/Code not Available		
1049	2009				CT257	Implantable cardiac pacemakers			A27	Appropriate Device Problem Term/Code not Available		

1050	2008				CT257	Implantable cardiac pacemakers			A27	Appropriate Device Problem Term/Code not Available		
1051	2007		47270	Cardiac resynchronization therapy implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	F1905	Device Revision or Replacement	A27	Appropriate Device Problem Term/Code not Available		
1052	2010				CT257	Implantable cardiac pacemakers			A27	Appropriate Device Problem Term/Code not Available		
1053	2010				CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A27	Appropriate Device Problem Term/Code not Available		
1054	2008				CT257	Implantable cardiac pacemakers			A27	Appropriate Device Problem Term/Code not Available		
1055	2006				CT257	Implantable cardiac pacemakers	F1905	Device Revision or Replacement	A27	Appropriate Device Problem Term/Code not Available		
1056	2006				CT257	Implantable cardiac pacemakers			A27	Appropriate Device Problem Term/Code not Available		
1057	2007				CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A27	Appropriate Device Problem Term/Code not Available		
1058	2010				CT257	Implantable cardiac pacemakers	F1905	Device Revision or Replacement	A27	Appropriate Device Problem Term/Code not Available		
1059	2010				CT257	Implantable cardiac pacemakers			A26	Insufficient Device Problem Information		
1060	2009				CT257	Implantable cardiac pacemakers			A27	Appropriate Device Problem Term/Code not Available		
1061	2010				CT257	Implantable cardiac pacemakers			A26	Insufficient Device Problem Information		
1062	2008				CT257	Implantable cardiac pacemakers			A27	Appropriate Device Problem Term/Code not Available		
1063	2009				CT257	Implantable cardiac pacemakers			A27	Appropriate Device Problem Term/Code not Available		
1064	2008				CT257	Implantable cardiac pacemakers			A27	Appropriate Device Problem Term/Code not Available		
1065	2006				CT257	Implantable cardiac pacemakers	F1905	Device Revision or Replacement	A27	Appropriate Device Problem Term/Code not Available		
1066	2008	All other reportable incidents			CT257	Implantable cardiac pacemakers			A27	Appropriate Device Problem Term/Code not Available		
1067	2007		47270	Cardiac resynchronization therapy implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	F1905	Device Revision or Replacement	A27	Appropriate Device Problem Term/Code not Available		
1068	2006		35223	Endocardial/interventricular septal pacing lead	CT1684	Cardiac pacemakers and associated devices	F1905	Device Revision or Replacement	A27	Appropriate Device Problem Term/Code not Available		
1069	2010				CT257	Implantable cardiac pacemakers			A27	Appropriate Device Problem Term/Code not Available		
1070	2009				CT257	Implantable cardiac pacemakers			A27	Appropriate Device Problem Term/Code not Available		
1071	2010				CT257	Implantable cardiac pacemakers			A27	Appropriate Device Problem Term/Code not Available		
1072	2008		47270	Cardiac resynchronization therapy implantable defibrillator	CT2866	Implantable cardiac electrical stimulators			A27	Appropriate Device Problem Term/Code not Available		
1073	2006				CT257	Implantable cardiac pacemakers	F1905	Device Revision or Replacement	A27	Appropriate Device Problem Term/Code not Available		
1074	2008		47270	Cardiac resynchronization therapy implantable defibrillator	CT2866	Implantable cardiac electrical stimulators			A27	Appropriate Device Problem Term/Code not Available		
1075	2007				CT257	Implantable cardiac pacemakers			A27	Appropriate Device Problem Term/Code not Available		
1076	2010				CT257	Implantable cardiac pacemakers	F1905	Device Revision or Replacement	A27	Appropriate Device Problem Term/Code not Available		
1077	2010				CT257	Implantable cardiac pacemakers			A27	Appropriate Device Problem Term/Code not Available		
1078	2007		47270	Cardiac resynchronization therapy implantable defibrillator	CT2866	Implantable cardiac electrical stimulators			A27	Appropriate Device Problem Term/Code not Available		
1079	2007		47270	Cardiac resynchronization therapy implantable defibrillator	CT2866	Implantable cardiac electrical stimulators			A27	Appropriate Device Problem Term/Code not Available		
1080	2010				CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A27	Appropriate Device Problem Term/Code not Available		
1081	2008		47270	Cardiac resynchronization therapy implantable defibrillator	CT2866	Implantable cardiac electrical stimulators			A27	Appropriate Device Problem Term/Code not Available		
1082	2009				CT257	Implantable cardiac pacemakers			A26	Insufficient Device Problem Information		
1083	2010	All other reportable incidents			CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A27	Appropriate Device Problem Term/Code not Available		
1084	2006				CT257	Implantable cardiac pacemakers	F1905	Device Revision or Replacement	A27	Appropriate Device Problem Term/Code not Available		
1085	2010				CT257	Implantable cardiac pacemakers			A27	Appropriate Device Problem Term/Code not Available		
1086	2010		47270	Cardiac resynchronization therapy implantable defibrillator	CT2866	Implantable cardiac electrical stimulators			A27	Appropriate Device Problem Term/Code not Available		
1087	2006				CT257	Implantable cardiac pacemakers			A27	Appropriate Device Problem Term/Code not Available		
1088	2008		47270	Cardiac resynchronization therapy implantable defibrillator	CT2866	Implantable cardiac electrical stimulators			A27	Appropriate Device Problem Term/Code not Available		
1089	2009		47270	Cardiac resynchronization therapy implantable defibrillator	CT2866	Implantable cardiac electrical stimulators			A27	Appropriate Device Problem Term/Code not Available		
1090	2010				CT257	Implantable cardiac pacemakers			A27	Appropriate Device Problem Term/Code not Available		
1091	2006				CT257	Implantable cardiac pacemakers			A27	Appropriate Device Problem Term/Code not Available		
1092	2008				CT257	Implantable cardiac pacemakers			A27	Appropriate Device Problem Term/Code not Available		
1093	2006				CT257	Implantable cardiac pacemakers			A27	Appropriate Device Problem Term/Code not Available		
1094	2007				CT257	Implantable cardiac pacemakers			A27	Appropriate Device Problem Term/Code not Available		
1095	2008		47270	Cardiac resynchronization therapy implantable defibrillator	CT2866	Implantable cardiac electrical stimulators			A26	Insufficient Device Problem Information		
1096	2007				CT257	Implantable cardiac pacemakers			A27	Appropriate Device Problem Term/Code not Available		
1097	2007				CT257	Implantable cardiac pacemakers			A27	Appropriate Device Problem Term/Code not Available		
1098	2009		35223	Endocardial/interventricular septal pacing lead	CT1684	Cardiac pacemakers and associated devices			A26	Insufficient Device Problem Information		
1099	2006				CT257	Implantable cardiac pacemakers			A27	Appropriate Device Problem Term/Code not Available		
1100	2006				CT257	Implantable cardiac pacemakers			A27	Appropriate Device Problem Term/Code not Available		

1101	2007				CT257	Implantable cardiac pacemakers			A27	Appropriate Device Problem Term/Code not Available		
1102	2009				CT257	Implantable cardiac pacemakers			A27	Appropriate Device Problem Term/Code not Available		
1103	2009		35223	Endocardial/interventricular septal pacing lead	CT1684	Cardiac pacemakers and associated devices			A27	Appropriate Device Problem Term/Code not Available		
1104	2009		47270	Cardiac resynchronization therapy implantable defibrillator	CT2866	Implantable cardiac electrical stimulators			A27	Appropriate Device Problem Term/Code not Available		
1105	2010				CT257	Implantable cardiac pacemakers			A27	Appropriate Device Problem Term/Code not Available		
1106	2008		35223	Endocardial/interventricular septal pacing lead	CT1684	Cardiac pacemakers and associated devices			A27	Appropriate Device Problem Term/Code not Available		
1107	2008		47270	Cardiac resynchronization therapy implantable defibrillator	CT2866	Implantable cardiac electrical stimulators			A27	Appropriate Device Problem Term/Code not Available		
1108	2008				CT257	Implantable cardiac pacemakers			A27	Appropriate Device Problem Term/Code not Available		
1109	2009				CT257	Implantable cardiac pacemakers			A27	Appropriate Device Problem Term/Code not Available		
1110	2010		35223	Endocardial/interventricular septal pacing lead	CT1684	Cardiac pacemakers and associated devices			A26	Insufficient Device Problem Information		
1111	2009				CT257	Implantable cardiac pacemakers			A27	Appropriate Device Problem Term/Code not Available		
1112	2008				CT257	Implantable cardiac pacemakers			A27	Appropriate Device Problem Term/Code not Available		
1113	2007		35223	Endocardial/interventricular septal pacing lead	CT1684	Cardiac pacemakers and associated devices			A27	Appropriate Device Problem Term/Code not Available		
1114	2007		35223	Endocardial/interventricular septal pacing lead	CT1684	Cardiac pacemakers and associated devices			A27	Appropriate Device Problem Term/Code not Available		
1115	2008				CT257	Implantable cardiac pacemakers			A27	Appropriate Device Problem Term/Code not Available		
1116	2006				CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A27	Appropriate Device Problem Term/Code not Available		
1117	2008				CT257	Implantable cardiac pacemakers			A27	Appropriate Device Problem Term/Code not Available		
1118	2007				CT257	Implantable cardiac pacemakers			A27	Appropriate Device Problem Term/Code not Available		
1119	2006				CT257	Implantable cardiac pacemakers	F1905	Device Revision or Replacement	A27	Appropriate Device Problem Term/Code not Available		
1120	2008				CT257	Implantable cardiac pacemakers			A27	Appropriate Device Problem Term/Code not Available		
1121	2008				CT257	Implantable cardiac pacemakers			A27	Appropriate Device Problem Term/Code not Available		
1122	2009				CT257	Implantable cardiac pacemakers			A27	Appropriate Device Problem Term/Code not Available		
1123	2009		47270	Cardiac resynchronization therapy implantable defibrillator	CT2866	Implantable cardiac electrical stimulators			A27	Appropriate Device Problem Term/Code not Available		
1124	2007				CT257	Implantable cardiac pacemakers			A27	Appropriate Device Problem Term/Code not Available		
1125	2007				CT257	Implantable cardiac pacemakers			A27	Appropriate Device Problem Term/Code not Available		
1126	2010		35223	Endocardial/interventricular septal pacing lead	CT1684	Cardiac pacemakers and associated devices			A27	Appropriate Device Problem Term/Code not Available		
1127	2006				CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A27	Appropriate Device Problem Term/Code not Available		
1128	2008				CT257	Implantable cardiac pacemakers			A27	Appropriate Device Problem Term/Code not Available		
1129	2009	All other reportable incidents			CT257	Implantable cardiac pacemakers			A27	Appropriate Device Problem Term/Code not Available		
1130	2009		47270	Cardiac resynchronization therapy implantable defibrillator	CT2866	Implantable cardiac electrical stimulators			A27	Appropriate Device Problem Term/Code not Available		
1131	2006				CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A27	Appropriate Device Problem Term/Code not Available		
1132	2010		35223	Endocardial/interventricular septal pacing lead	CT1684	Cardiac pacemakers and associated devices			A26	Insufficient Device Problem Information		
1133	2008		47270	Cardiac resynchronization therapy implantable defibrillator	CT2866	Implantable cardiac electrical stimulators			A27	Appropriate Device Problem Term/Code not Available		
1134	2010				CT257	Implantable cardiac pacemakers			A27	Appropriate Device Problem Term/Code not Available		
1135	2009				CT257	Implantable cardiac pacemakers			A27	Appropriate Device Problem Term/Code not Available		
1136	2006				CT257	Implantable cardiac pacemakers	F1905	Device Revision or Replacement	A27	Appropriate Device Problem Term/Code not Available		
1137	2010				CT257	Implantable cardiac pacemakers			A27	Appropriate Device Problem Term/Code not Available		
1138	2010				CT257	Implantable cardiac pacemakers			A27	Appropriate Device Problem Term/Code not Available		
1139	2009	All other reportable incidents			CT257	Implantable cardiac pacemakers			A27	Appropriate Device Problem Term/Code not Available		
1140	2006		35223	Endocardial/interventricular septal pacing lead	CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A27	Appropriate Device Problem Term/Code not Available		
1141	2008				CT257	Implantable cardiac pacemakers			A27	Appropriate Device Problem Term/Code not Available		
1142	2006		47270	Cardiac resynchronization therapy implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	F1905	Device Revision or Replacement	A27	Appropriate Device Problem Term/Code not Available		
1143	2008				CT257	Implantable cardiac pacemakers			A27	Appropriate Device Problem Term/Code not Available		
1144	2009	All other reportable incidents			CT257	Implantable cardiac pacemakers			A27	Appropriate Device Problem Term/Code not Available	23/07/2009	
1145	2007		35223	Endocardial/interventricular septal pacing lead	CT1684	Cardiac pacemakers and associated devices			A27	Appropriate Device Problem Term/Code not Available		
1146	2007				CT257	Implantable cardiac pacemakers			A27	Appropriate Device Problem Term/Code not Available		
1147	2007				CT257	Implantable cardiac pacemakers			A27	Appropriate Device Problem Term/Code not Available		
1148	2008				CT257	Implantable cardiac pacemakers			A27	Appropriate Device Problem Term/Code not Available		
1149	2009				CT257	Implantable cardiac pacemakers			A27	Appropriate Device Problem Term/Code not Available		
1150	2008				CT257	Implantable cardiac pacemakers			A27	Appropriate Device Problem Term/Code not Available		

1151	2008	All other reportable incidents			CT257	Implantable cardiac pacemakers			A27	Appropriate Device Problem Term/Code not Available		
1152	2007				CT257	Implantable cardiac pacemakers			A27	Appropriate Device Problem Term/Code not Available		
1153	2006				CT257	Implantable cardiac pacemakers	F1905	Device Revision or Replacement	A27	Appropriate Device Problem Term/Code not Available		
1154	2008		47270	Cardiac resynchronization therapy implantable defibrillator	CT2866	Implantable cardiac electrical stimulators			A27	Appropriate Device Problem Term/Code not Available		
1155	2008				CT257	Implantable cardiac pacemakers			A27	Appropriate Device Problem Term/Code not Available		
1156	2010				CT257	Implantable cardiac pacemakers			A26	Insufficient Device Problem Information		
1157	2007				CT257	Implantable cardiac pacemakers			A27	Appropriate Device Problem Term/Code not Available		
1158	2010				CT257	Implantable cardiac pacemakers			A27	Appropriate Device Problem Term/Code not Available		
1159	2008				CT257	Implantable cardiac pacemakers			A27	Appropriate Device Problem Term/Code not Available		
1160	2006				CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A27	Appropriate Device Problem Term/Code not Available		
1161	2010				CT257	Implantable cardiac pacemakers			A27	Appropriate Device Problem Term/Code not Available		
1162	2009				CT257	Implantable cardiac pacemakers			A27	Appropriate Device Problem Term/Code not Available		
1163	2008		35223	Endocardial/interventricular septal pacing lead	CT1684	Cardiac pacemakers and associated devices			A27	Appropriate Device Problem Term/Code not Available		
1164	2007				CT257	Implantable cardiac pacemakers			A27	Appropriate Device Problem Term/Code not Available		
1165	2008				CT257	Implantable cardiac pacemakers			A26	Insufficient Device Problem Information		
1166	2007		47270	Cardiac resynchronization therapy implantable defibrillator	CT2866	Implantable cardiac electrical stimulators			A27	Appropriate Device Problem Term/Code not Available		
1167	2009				CT257	Implantable cardiac pacemakers			A27	Appropriate Device Problem Term/Code not Available		
1168	2007				CT257	Implantable cardiac pacemakers			A27	Appropriate Device Problem Term/Code not Available		
1169	2007		47270	Cardiac resynchronization therapy implantable defibrillator	CT2866	Implantable cardiac electrical stimulators			A27	Appropriate Device Problem Term/Code not Available		
1170	2010				CT1684	Cardiac pacemakers and associated devices			A27	Appropriate Device Problem Term/Code not Available		
1171	2006		35223	Endocardial/interventricular septal pacing lead	CT1684	Cardiac pacemakers and associated devices	F1905	Device Revision or Replacement	A27	Appropriate Device Problem Term/Code not Available		
1172	2008	All other reportable incidents			CT257	Implantable cardiac pacemakers			A27	Appropriate Device Problem Term/Code not Available		
1173	2009				CT257	Implantable cardiac pacemakers			A27	Appropriate Device Problem Term/Code not Available		
1174	2009				CT257	Implantable cardiac pacemakers			A26	Insufficient Device Problem Information		
1175	2008	All other reportable incidents			CT257	Implantable cardiac pacemakers			A27	Appropriate Device Problem Term/Code not Available		
1176	2010		47270	Cardiac resynchronization therapy implantable defibrillator	CT2866	Implantable cardiac electrical stimulators			A26	Insufficient Device Problem Information		
1177	2010				CT257	Implantable cardiac pacemakers			A27	Appropriate Device Problem Term/Code not Available		
1178	2010				CT257	Implantable cardiac pacemakers			A27	Appropriate Device Problem Term/Code not Available		
1179	2006				CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A27	Appropriate Device Problem Term/Code not Available		
1180	2006				CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A27	Appropriate Device Problem Term/Code not Available		
1181	2008				CT257	Implantable cardiac pacemakers			A27	Appropriate Device Problem Term/Code not Available		
1182	2007		47270	Cardiac resynchronization therapy implantable defibrillator	CT2866	Implantable cardiac electrical stimulators			A27	Appropriate Device Problem Term/Code not Available		
1183	2010				CT257	Implantable cardiac pacemakers			A27	Appropriate Device Problem Term/Code not Available		
1184	2006		47270	Cardiac resynchronization therapy implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A27	Appropriate Device Problem Term/Code not Available		
1185	2007				CT257	Implantable cardiac pacemakers			A27	Appropriate Device Problem Term/Code not Available		
1186	2007		35223	Endocardial/interventricular septal pacing lead	CT1684	Cardiac pacemakers and associated devices			A27	Appropriate Device Problem Term/Code not Available		
1187	2006		47270	Cardiac resynchronization therapy implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A27	Appropriate Device Problem Term/Code not Available		
1188	2006		47270	Cardiac resynchronization therapy implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A27	Appropriate Device Problem Term/Code not Available		
1189	2010		47270	Cardiac resynchronization therapy implantable defibrillator	CT2866	Implantable cardiac electrical stimulators			A26	Insufficient Device Problem Information		
1190	2008	All other reportable incidents			CT257	Implantable cardiac pacemakers			A27	Appropriate Device Problem Term/Code not Available		
1191	2008				CT257	Implantable cardiac pacemakers			A27	Appropriate Device Problem Term/Code not Available		
1192	2007		35039	Epicardial pacing lead	CT1684	Cardiac pacemakers and associated devices			A27	Appropriate Device Problem Term/Code not Available		
1193	2010		47270	Cardiac resynchronization therapy implantable defibrillator	CT2866	Implantable cardiac electrical stimulators			A27	Appropriate Device Problem Term/Code not Available		
1194	2010		47270	Cardiac resynchronization therapy implantable defibrillator	CT2866	Implantable cardiac electrical stimulators			A27	Appropriate Device Problem Term/Code not Available		
1195	2008				CT257	Implantable cardiac pacemakers			A27	Appropriate Device Problem Term/Code not Available		
1196	2006		35223	Endocardial/interventricular septal pacing lead	CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A27	Appropriate Device Problem Term/Code not Available		
1197	2009				CT257	Implantable cardiac pacemakers			A27	Appropriate Device Problem Term/Code not Available		
1198	2006		47270	Cardiac resynchronization therapy implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A27	Appropriate Device Problem Term/Code not Available		
1199	2008		35223	Endocardial/interventricular septal pacing lead	CT1684	Cardiac pacemakers and associated devices			A27	Appropriate Device Problem Term/Code not Available		
1200	2007				CT257	Implantable cardiac pacemakers			A27	Appropriate Device Problem Term/Code not Available		
1201	2008		35223	Endocardial/interventricular septal pacing lead	CT1684	Cardiac pacemakers and associated devices			A27	Appropriate Device Problem Term/Code not Available		

1202	2008				CT257	Implantable cardiac pacemakers			A27	Appropriate Device Problem Term/Code not Available		
1203	2009		35223	Endocardial/interventricular septal pacing lead	CT1684	Cardiac pacemakers and associated devices			A27	Appropriate Device Problem Term/Code not Available		
1204	2006	All other reportable incidents			CT257	Implantable cardiac pacemakers			A27	Appropriate Device Problem Term/Code not Available		
1205	2006	All other reportable incidents			CT257	Implantable cardiac pacemakers			A27	Appropriate Device Problem Term/Code not Available		
1206	2006				CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A27	Appropriate Device Problem Term/Code not Available		
1207	2009				CT257	Implantable cardiac pacemakers			A27	Appropriate Device Problem Term/Code not Available		
1208	2006		35223	Endocardial/interventricular septal pacing lead	CT1684	Cardiac pacemakers and associated devices			A27	Appropriate Device Problem Term/Code not Available		
1209	2007		35223	Endocardial/interventricular septal pacing lead	CT1684	Cardiac pacemakers and associated devices			A27	Appropriate Device Problem Term/Code not Available		
1210	2007		47270	Cardiac resynchronization therapy implantable defibrillator	CT2866	Implantable cardiac electrical stimulators			A27	Appropriate Device Problem Term/Code not Available		
1211	2007				CT257	Implantable cardiac pacemakers			A27	Appropriate Device Problem Term/Code not Available		
1212	2010				CT257	Implantable cardiac pacemakers			A27	Appropriate Device Problem Term/Code not Available		
1213	2008				CT257	Implantable cardiac pacemakers			A27	Appropriate Device Problem Term/Code not Available		
1214	2010				CT257	Implantable cardiac pacemakers			A26	Insufficient Device Problem Information		
1215	2007		47270	Cardiac resynchronization therapy implantable defibrillator	CT2866	Implantable cardiac electrical stimulators			A27	Appropriate Device Problem Term/Code not Available		
1216	2008				CT257	Implantable cardiac pacemakers			A27	Appropriate Device Problem Term/Code not Available		
1217	2008		47270	Cardiac resynchronization therapy implantable defibrillator	CT2866	Implantable cardiac electrical stimulators			A27	Appropriate Device Problem Term/Code not Available		
1218	2010				CT257	Implantable cardiac pacemakers			A27	Appropriate Device Problem Term/Code not Available		
1219	2007				CT257	Implantable cardiac pacemakers			A27	Appropriate Device Problem Term/Code not Available		
1220	2009				CT257	Implantable cardiac pacemakers			A26	Insufficient Device Problem Information		
1221	2009				CT257	Implantable cardiac pacemakers			A27	Appropriate Device Problem Term/Code not Available		
1222	2007		35223	Endocardial/interventricular septal pacing lead	CT1684	Cardiac pacemakers and associated devices			A27	Appropriate Device Problem Term/Code not Available		
1223	2009				CT257	Implantable cardiac pacemakers			A27	Appropriate Device Problem Term/Code not Available		
1224	2009				CT257	Implantable cardiac pacemakers			A27	Appropriate Device Problem Term/Code not Available		
1225	2010		47270	Cardiac resynchronization therapy implantable defibrillator	CT2866	Implantable cardiac electrical stimulators			A27	Appropriate Device Problem Term/Code not Available		
1226	2010		60910	Coronary venous pacing lead	CT1684	Cardiac pacemakers and associated devices			A27	Appropriate Device Problem Term/Code not Available		
1227	2010		35223	Endocardial/interventricular septal pacing lead	CT1684	Cardiac pacemakers and associated devices			A27	Appropriate Device Problem Term/Code not Available		
1228	2010				CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A27	Appropriate Device Problem Term/Code not Available		
1229	2008				CT257	Implantable cardiac pacemakers			A27	Appropriate Device Problem Term/Code not Available		
1230	2010				CT257	Implantable cardiac pacemakers			A26	Insufficient Device Problem Information		
1231	2009				CT257	Implantable cardiac pacemakers			A27	Appropriate Device Problem Term/Code not Available		
1232	2008				CT257	Implantable cardiac pacemakers			A27	Appropriate Device Problem Term/Code not Available		
1233	2007		47270	Cardiac resynchronization therapy implantable defibrillator	CT2866	Implantable cardiac electrical stimulators			A27	Appropriate Device Problem Term/Code not Available		
1234	2006		35223	Endocardial/interventricular septal pacing lead	CT1684	Cardiac pacemakers and associated devices	F1905	Device Revision or Replacement	A27	Appropriate Device Problem Term/Code not Available		
1235	2008				CT257	Implantable cardiac pacemakers			A27	Appropriate Device Problem Term/Code not Available		
1236	2008		47270	Cardiac resynchronization therapy implantable defibrillator	CT2866	Implantable cardiac electrical stimulators			A27	Appropriate Device Problem Term/Code not Available		
1237	2009	All other reportable incidents			CT257	Implantable cardiac pacemakers			A27	Appropriate Device Problem Term/Code not Available		
1238	2009				CT257	Implantable cardiac pacemakers			A27	Appropriate Device Problem Term/Code not Available		
1239	2008		35223	Endocardial/interventricular septal pacing lead	CT1684	Cardiac pacemakers and associated devices			A27	Appropriate Device Problem Term/Code not Available		
1240	2008				CT257	Implantable cardiac pacemakers			A27	Appropriate Device Problem Term/Code not Available		
1241	2008				CT257	Implantable cardiac pacemakers			A27	Appropriate Device Problem Term/Code not Available		
1242	2008		35223	Endocardial/interventricular septal pacing lead	CT1684	Cardiac pacemakers and associated devices			A27	Appropriate Device Problem Term/Code not Available		
1243	2009	All other reportable incidents			CT257	Implantable cardiac pacemakers			A27	Appropriate Device Problem Term/Code not Available		
1244	2009	All other reportable incidents			CT257	Implantable cardiac pacemakers			A27	Appropriate Device Problem Term/Code not Available		
1245	2009				CT257	Implantable cardiac pacemakers			A27	Appropriate Device Problem Term/Code not Available		
1246	2007		35223	Endocardial/interventricular septal pacing lead	CT1684	Cardiac pacemakers and associated devices			A27	Appropriate Device Problem Term/Code not Available		
1247	2009	All other reportable incidents			CT257	Implantable cardiac pacemakers			A27	Appropriate Device Problem Term/Code not Available		
1248	2006				CT257	Implantable cardiac pacemakers	F1905	Device Revision or Replacement	A27	Appropriate Device Problem Term/Code not Available		
1249	2006				CT257	Implantable cardiac pacemakers	F1905	Device Revision or Replacement	A27	Appropriate Device Problem Term/Code not Available		
1250	2008				CT257	Implantable cardiac pacemakers			A27	Appropriate Device Problem Term/Code not Available		
1251	2010		47263	Cardiac resynchronization therapy implantable pacemaker	CT1684	Cardiac pacemakers and associated devices			A27	Appropriate Device Problem Term/Code not Available		
					CT1685	Cardiac pacemakers			A27	Appropriate Device Problem Term/Code not Available		

					CT257	Implantable cardiac pacemakers			A27	Appropriate Device Problem Term/Code not Available		
					CT2866	Implantable cardiac electrical stimulators			A27	Appropriate Device Problem Term/Code not Available		
1252	2010				CT257	Implantable cardiac pacemakers			A27	Appropriate Device Problem Term/Code not Available		
1253	2010				CT257	Implantable cardiac pacemakers			A27	Appropriate Device Problem Term/Code not Available		
1254	2010				CT257	Implantable cardiac pacemakers			A26	Insufficient Device Problem Information		
1255	2010				CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A27	Appropriate Device Problem Term/Code not Available		
1256	2008		35223	Endocardial/interventricular septal pacing lead	CT1684	Cardiac pacemakers and associated devices			A27	Appropriate Device Problem Term/Code not Available		
1257	2006				CT257	Implantable cardiac pacemakers	F1905	Device Revision or Replacement	A27	Appropriate Device Problem Term/Code not Available		
1258	2008		35223	Endocardial/interventricular septal pacing lead	CT1684	Cardiac pacemakers and associated devices			A27	Appropriate Device Problem Term/Code not Available		
1259	2008		47270	Cardiac resynchronization therapy implantable defibrillator	CT2866	Implantable cardiac electrical stimulators			A27	Appropriate Device Problem Term/Code not Available		
1260	2009				CT257	Implantable cardiac pacemakers			A27	Appropriate Device Problem Term/Code not Available		
1261	2009	All other reportable incidents			CT257	Implantable cardiac pacemakers			A26	Insufficient Device Problem Information		
1262	2009		35223	Endocardial/interventricular septal pacing lead	CT1684	Cardiac pacemakers and associated devices			A27	Appropriate Device Problem Term/Code not Available		
1263	2007		47270	Cardiac resynchronization therapy implantable defibrillator	CT2866	Implantable cardiac electrical stimulators			A27	Appropriate Device Problem Term/Code not Available		
1264	2008	All other reportable incidents			CT257	Implantable cardiac pacemakers			A27	Appropriate Device Problem Term/Code not Available		
1265	2009				CT257	Implantable cardiac pacemakers			A27	Appropriate Device Problem Term/Code not Available		
1266	2009		47270	Cardiac resynchronization therapy implantable defibrillator	CT2866	Implantable cardiac electrical stimulators			A27	Appropriate Device Problem Term/Code not Available		
1267	2009				CT257	Implantable cardiac pacemakers			A27	Appropriate Device Problem Term/Code not Available		
1268	2009		61438	Cardiac pulse generator reader, home-use	CT1684	Cardiac pacemakers and associated devices			A27	Appropriate Device Problem Term/Code not Available		
1269	2010				CT257	Implantable cardiac pacemakers			A27	Appropriate Device Problem Term/Code not Available		
1270	2008				CT257	Implantable cardiac pacemakers			A27	Appropriate Device Problem Term/Code not Available		
1271	2006				CT257	Implantable cardiac pacemakers	F1905	Device Revision or Replacement	A27	Appropriate Device Problem Term/Code not Available		
1272	2007		47270	Cardiac resynchronization therapy implantable defibrillator	CT2866	Implantable cardiac electrical stimulators			A27	Appropriate Device Problem Term/Code not Available		
1273	2009				CT257	Implantable cardiac pacemakers			A27	Appropriate Device Problem Term/Code not Available		
1274	2010				CT257	Implantable cardiac pacemakers			A27	Appropriate Device Problem Term/Code not Available		
1275	2009		47270	Cardiac resynchronization therapy implantable defibrillator	CT2866	Implantable cardiac electrical stimulators			A27	Appropriate Device Problem Term/Code not Available		
1276	2009		47270	Cardiac resynchronization therapy implantable defibrillator	CT2866	Implantable cardiac electrical stimulators			A27	Appropriate Device Problem Term/Code not Available		
1277	2008				CT257	Implantable cardiac pacemakers			A27	Appropriate Device Problem Term/Code not Available		
1278	2009		47270	Cardiac resynchronization therapy implantable defibrillator	CT2866	Implantable cardiac electrical stimulators			A27	Appropriate Device Problem Term/Code not Available		
1279	2010				CT257	Implantable cardiac pacemakers			A27	Appropriate Device Problem Term/Code not Available		
1280	2006				CT257	Implantable cardiac pacemakers			A27	Appropriate Device Problem Term/Code not Available		
1281	2010				CT257	Implantable cardiac pacemakers			A27	Appropriate Device Problem Term/Code not Available		
1282	2008		47270	Cardiac resynchronization therapy implantable defibrillator	CT2866	Implantable cardiac electrical stimulators			A27	Appropriate Device Problem Term/Code not Available		
1283	2009				CT257	Implantable cardiac pacemakers			A27	Appropriate Device Problem Term/Code not Available		
1284	2010		47270	Cardiac resynchronization therapy implantable defibrillator	CT2866	Implantable cardiac electrical stimulators			A26	Insufficient Device Problem Information		
1285	2007	All other reportable incidents			CT257	Implantable cardiac pacemakers			A27	Appropriate Device Problem Term/Code not Available		
1286	2008				CT257	Implantable cardiac pacemakers			A27	Appropriate Device Problem Term/Code not Available		
1287	2008		61438	Cardiac pulse generator reader, home-use	CT1684	Cardiac pacemakers and associated devices			A27	Appropriate Device Problem Term/Code not Available		
1288	2010		47270	Cardiac resynchronization therapy implantable defibrillator	CT2866	Implantable cardiac electrical stimulators			A27	Appropriate Device Problem Term/Code not Available		
1289	2009	All other reportable incidents			CT257	Implantable cardiac pacemakers			A27	Appropriate Device Problem Term/Code not Available		
1290	2009		47270	Cardiac resynchronization therapy implantable defibrillator	CT2866	Implantable cardiac electrical stimulators			A27	Appropriate Device Problem Term/Code not Available		
1291	2010				CT257	Implantable cardiac pacemakers			A27	Appropriate Device Problem Term/Code not Available		
1292	2007		47270	Cardiac resynchronization therapy implantable defibrillator	CT2866	Implantable cardiac electrical stimulators			A27	Appropriate Device Problem Term/Code not Available		
1293	2006				CT257	Implantable cardiac pacemakers			A27	Appropriate Device Problem Term/Code not Available		
1294	2010				CT257	Implantable cardiac pacemakers			A27	Appropriate Device Problem Term/Code not Available		
1295	2010				CT257	Implantable cardiac pacemakers			A27	Appropriate Device Problem Term/Code not Available		
1296	2009		47270	Cardiac resynchronization therapy implantable defibrillator	CT2866	Implantable cardiac electrical stimulators			A27	Appropriate Device Problem Term/Code not Available		
1297	2009				CT257	Implantable cardiac pacemakers			A27	Appropriate Device Problem Term/Code not Available	12/10/2009	
1298	2007				CT257	Implantable cardiac pacemakers			A27	Appropriate Device Problem Term/Code not Available		
1299	2009				CT257	Implantable cardiac pacemakers			A27	Appropriate Device Problem Term/Code not Available		

1300	2007				CT257	Implantable cardiac pacemakers			A27	Appropriate Device Problem Term/Code not Available		
1301	2007				CT257	Implantable cardiac pacemakers			A27	Appropriate Device Problem Term/Code not Available		
1302	2010		47270	Cardiac resynchronization therapy implantable defibrillator	CT2866	Implantable cardiac electrical stimulators			A27	Appropriate Device Problem Term/Code not Available		
1303	2010				CT257	Implantable cardiac pacemakers			A27	Appropriate Device Problem Term/Code not Available		
1304	2007		35223	Endocardial/interventricular septal pacing lead	CT1684	Cardiac pacemakers and associated devices			A27	Appropriate Device Problem Term/Code not Available		
1305	2009				CT257	Implantable cardiac pacemakers			A27	Appropriate Device Problem Term/Code not Available		
1306	2007				CT257	Implantable cardiac pacemakers			A26	Insufficient Device Problem Information		
1307	2007				CT257	Implantable cardiac pacemakers			A27	Appropriate Device Problem Term/Code not Available		
1308	2006		47270	Cardiac resynchronization therapy implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	F1905	Device Revision or Replacement	A27	Appropriate Device Problem Term/Code not Available		
1309	2007				CT257	Implantable cardiac pacemakers			A27	Appropriate Device Problem Term/Code not Available		
1310	2006				CT257	Implantable cardiac pacemakers	F1905	Device Revision or Replacement	A27	Appropriate Device Problem Term/Code not Available		
1311	2010				CT257	Implantable cardiac pacemakers			A27	Appropriate Device Problem Term/Code not Available		
1312	2007				CT257	Implantable cardiac pacemakers			A27	Appropriate Device Problem Term/Code not Available		
1313	2009				CT257	Implantable cardiac pacemakers			A27	Appropriate Device Problem Term/Code not Available		
1314	2009				CT257	Implantable cardiac pacemakers			A27	Appropriate Device Problem Term/Code not Available		
1315	2009				CT257	Implantable cardiac pacemakers			A27	Appropriate Device Problem Term/Code not Available		
1316	2007		47270	Cardiac resynchronization therapy implantable defibrillator	CT2866	Implantable cardiac electrical stimulators			A27	Appropriate Device Problem Term/Code not Available		
1317	2008		47270	Cardiac resynchronization therapy implantable defibrillator	CT2866	Implantable cardiac electrical stimulators			A27	Appropriate Device Problem Term/Code not Available		
1318	2007		47270	Cardiac resynchronization therapy implantable defibrillator	CT2866	Implantable cardiac electrical stimulators			A27	Appropriate Device Problem Term/Code not Available		
1319	2008	All other reportable incidents			CT257	Implantable cardiac pacemakers			A27	Appropriate Device Problem Term/Code not Available		
1320	2010				CT257	Implantable cardiac pacemakers			A27	Appropriate Device Problem Term/Code not Available		
1321	2008				CT257	Implantable cardiac pacemakers			A27	Appropriate Device Problem Term/Code not Available		
1322	2009	All other reportable incidents			CT257	Implantable cardiac pacemakers			A27	Appropriate Device Problem Term/Code not Available		
1323	2006				CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A27	Appropriate Device Problem Term/Code not Available		
1324	2006				CT257	Implantable cardiac pacemakers			A27	Appropriate Device Problem Term/Code not Available		
1325	2010				CT257	Implantable cardiac pacemakers			A27	Appropriate Device Problem Term/Code not Available		
1326	2010		47270	Cardiac resynchronization therapy implantable defibrillator	CT2866	Implantable cardiac electrical stimulators			A26	Insufficient Device Problem Information		
1327	2008				CT257	Implantable cardiac pacemakers			A27	Appropriate Device Problem Term/Code not Available		
1328	2007		47270	Cardiac resynchronization therapy implantable defibrillator	CT2866	Implantable cardiac electrical stimulators			A27	Appropriate Device Problem Term/Code not Available		
1329	2008		35223	Endocardial/interventricular septal pacing lead	CT1684	Cardiac pacemakers and associated devices			A26	Insufficient Device Problem Information		
1330	2008				CT257	Implantable cardiac pacemakers			A27	Appropriate Device Problem Term/Code not Available		
1331	2008				CT257	Implantable cardiac pacemakers			A27	Appropriate Device Problem Term/Code not Available		
1332	2006				CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A27	Appropriate Device Problem Term/Code not Available		
1333	2006				CT257	Implantable cardiac pacemakers			A27	Appropriate Device Problem Term/Code not Available		
1334	2009				CT257	Implantable cardiac pacemakers			A27	Appropriate Device Problem Term/Code not Available		
1335	2010				CT257	Implantable cardiac pacemakers			A27	Appropriate Device Problem Term/Code not Available		
1336	2008				CT257	Implantable cardiac pacemakers			A27	Appropriate Device Problem Term/Code not Available		
1337	2008		35223	Endocardial/interventricular septal pacing lead	CT1684	Cardiac pacemakers and associated devices			A27	Appropriate Device Problem Term/Code not Available		
1338	2010		47270	Cardiac resynchronization therapy implantable defibrillator	CT2866	Implantable cardiac electrical stimulators			A27	Appropriate Device Problem Term/Code not Available		
1339	2008				CT257	Implantable cardiac pacemakers			A27	Appropriate Device Problem Term/Code not Available		
1340	2008				CT257	Implantable cardiac pacemakers			A27	Appropriate Device Problem Term/Code not Available		
1341	2010		47270	Cardiac resynchronization therapy implantable defibrillator	CT2866	Implantable cardiac electrical stimulators			A27	Appropriate Device Problem Term/Code not Available		
1342	2006				CT257	Implantable cardiac pacemakers	F1905	Device Revision or Replacement	A27	Appropriate Device Problem Term/Code not Available		
1343	2006				CT257	Implantable cardiac pacemakers	F1905	Device Revision or Replacement	A27	Appropriate Device Problem Term/Code not Available		
1344	2008		35223	Endocardial/interventricular septal pacing lead	CT1684	Cardiac pacemakers and associated devices			A27	Appropriate Device Problem Term/Code not Available		
1345	2008	All other reportable incidents			CT257	Implantable cardiac pacemakers			A27	Appropriate Device Problem Term/Code not Available		
1346	2010				CT257	Implantable cardiac pacemakers			A27	Appropriate Device Problem Term/Code not Available		
1347	2009	All other reportable incidents			CT257	Implantable cardiac pacemakers			A27	Appropriate Device Problem Term/Code not Available		
1348	2007				CT257	Implantable cardiac pacemakers			A27	Appropriate Device Problem Term/Code not Available		
1349	2006		35223	Endocardial/interventricular septal pacing lead	CT1684	Cardiac pacemakers and associated devices	F1905	Device Revision or Replacement	A27	Appropriate Device Problem Term/Code not Available		

1350	2008				CT257	Implantable cardiac pacemakers			A27	Appropriate Device Problem Term/Code not Available		
1351	2007				CT257	Implantable cardiac pacemakers			A27	Appropriate Device Problem Term/Code not Available		
1352	2006				CT257	Implantable cardiac pacemakers			A27	Appropriate Device Problem Term/Code not Available		
1353	2007				CT257	Implantable cardiac pacemakers			A27	Appropriate Device Problem Term/Code not Available		
1354	2010				CT257	Implantable cardiac pacemakers			A27	Appropriate Device Problem Term/Code not Available		
1355	2008		35223	Endocardial/interventricular septal pacing lead	CT1684	Cardiac pacemakers and associated devices			A27	Appropriate Device Problem Term/Code not Available		
1356	2006				CT257	Implantable cardiac pacemakers	F1905	Device Revision or Replacement	A27	Appropriate Device Problem Term/Code not Available		
1357	2009	All other reportable incidents			CT257	Implantable cardiac pacemakers			A27	Appropriate Device Problem Term/Code not Available		
1358	2009				CT257	Implantable cardiac pacemakers			A27	Appropriate Device Problem Term/Code not Available		
1359	2008	All other reportable incidents			CT257	Implantable cardiac pacemakers			A27	Appropriate Device Problem Term/Code not Available		
1360	2006				CT257	Implantable cardiac pacemakers	F1905	Device Revision or Replacement	A27	Appropriate Device Problem Term/Code not Available		
1361	2006				CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A27	Appropriate Device Problem Term/Code not Available		
1362	2010				CT257	Implantable cardiac pacemakers			A27	Appropriate Device Problem Term/Code not Available		
1363	2008				CT257	Implantable cardiac pacemakers			A27	Appropriate Device Problem Term/Code not Available		
1364	2008				CT257	Implantable cardiac pacemakers			A27	Appropriate Device Problem Term/Code not Available		
1365	2007				CT257	Implantable cardiac pacemakers			A27	Appropriate Device Problem Term/Code not Available		
1366	2007				CT257	Implantable cardiac pacemakers			A27	Appropriate Device Problem Term/Code not Available		
1367	2009		47270	Cardiac resynchronization therapy implantable defibrillator	CT2866	Implantable cardiac electrical stimulators			A27	Appropriate Device Problem Term/Code not Available		
1368	2008	Unanticipated serious deterioration in state of health			CT257	Implantable cardiac pacemakers			A27	Appropriate Device Problem Term/Code not Available		
1369	2007				CT257	Implantable cardiac pacemakers			A27	Appropriate Device Problem Term/Code not Available		
1370	2010				CT257	Implantable cardiac pacemakers			A27	Appropriate Device Problem Term/Code not Available		
1371	2006		60910	Coronary venous pacing lead	CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A27	Appropriate Device Problem Term/Code not Available		
1372	2010				CT257	Implantable cardiac pacemakers			A27	Appropriate Device Problem Term/Code not Available		
1373	2008				CT257	Implantable cardiac pacemakers			A26	Insufficient Device Problem Information		
1374	2008		47270	Cardiac resynchronization therapy implantable defibrillator	CT2866	Implantable cardiac electrical stimulators			A27	Appropriate Device Problem Term/Code not Available		
1375	2008				CT257	Implantable cardiac pacemakers			A27	Appropriate Device Problem Term/Code not Available		
1376	2006				CT257	Implantable cardiac pacemakers	F1905	Device Revision or Replacement	A27	Appropriate Device Problem Term/Code not Available		
1377	2007		47270	Cardiac resynchronization therapy implantable defibrillator	CT2866	Implantable cardiac electrical stimulators			A27	Appropriate Device Problem Term/Code not Available		
1378	2007		47270	Cardiac resynchronization therapy implantable defibrillator	CT2866	Implantable cardiac electrical stimulators			A27	Appropriate Device Problem Term/Code not Available		
1379	2007		47270	Cardiac resynchronization therapy implantable defibrillator	CT2866	Implantable cardiac electrical stimulators			A27	Appropriate Device Problem Term/Code not Available		
1380	2008		47270	Cardiac resynchronization therapy implantable defibrillator	CT2866	Implantable cardiac electrical stimulators			A27	Appropriate Device Problem Term/Code not Available		
1381	2010		47270	Cardiac resynchronization therapy implantable defibrillator	CT2866	Implantable cardiac electrical stimulators			A27	Appropriate Device Problem Term/Code not Available		
1382	2010		47270	Cardiac resynchronization therapy implantable defibrillator	CT2866	Implantable cardiac electrical stimulators			A27	Appropriate Device Problem Term/Code not Available		
1383	2009		35223	Endocardial/interventricular septal pacing lead	CT1684	Cardiac pacemakers and associated devices			A27	Appropriate Device Problem Term/Code not Available		
1384	2008		35223	Endocardial/interventricular septal pacing lead	CT1684	Cardiac pacemakers and associated devices			A27	Appropriate Device Problem Term/Code not Available		
1385	2007				CT257	Implantable cardiac pacemakers			A27	Appropriate Device Problem Term/Code not Available		
1386	2007				CT257	Implantable cardiac pacemakers			A27	Appropriate Device Problem Term/Code not Available		
1387	2008	All other reportable incidents			CT257	Implantable cardiac pacemakers			A27	Appropriate Device Problem Term/Code not Available		
1388	2006				CT257	Implantable cardiac pacemakers	F1905	Device Revision or Replacement	A27	Appropriate Device Problem Term/Code not Available		
1389	2009		35223	Endocardial/interventricular septal pacing lead	CT1684	Cardiac pacemakers and associated devices			A27	Appropriate Device Problem Term/Code not Available		
1390	2007		47270	Cardiac resynchronization therapy implantable defibrillator	CT2866	Implantable cardiac electrical stimulators			A27	Appropriate Device Problem Term/Code not Available		
1391	2007		47270	Cardiac resynchronization therapy implantable defibrillator	CT2866	Implantable cardiac electrical stimulators			A27	Appropriate Device Problem Term/Code not Available		
1392	2007		47270	Cardiac resynchronization therapy implantable defibrillator	CT2866	Implantable cardiac electrical stimulators			A27	Appropriate Device Problem Term/Code not Available		
1393	2008	All other reportable incidents			CT257	Implantable cardiac pacemakers			A27	Appropriate Device Problem Term/Code not Available		
1394	2008		47270	Cardiac resynchronization therapy implantable defibrillator	CT2866	Implantable cardiac electrical stimulators			A27	Appropriate Device Problem Term/Code not Available		
1395	2010				CT257	Implantable cardiac pacemakers			A27	Appropriate Device Problem Term/Code not Available		
1396	2008				CT257	Implantable cardiac pacemakers			A27	Appropriate Device Problem Term/Code not Available		
1397	2009				CT257	Implantable cardiac pacemakers			A26	Insufficient Device Problem Information		
1398	2010				CT257	Implantable cardiac pacemakers			A26	Insufficient Device Problem Information		
1399	2008				CT257	Implantable cardiac pacemakers			A27	Appropriate Device Problem Term/Code not Available		
1400	2010				CT257	Implantable cardiac pacemakers			A26	Insufficient Device Problem Information		

1401	2007				CT257	Implantable cardiac pacemakers			A27	Appropriate Device Problem Term/Code not Available		
1402	2009				CT257	Implantable cardiac pacemakers			A27	Appropriate Device Problem Term/Code not Available		
1403	2008				CT257	Implantable cardiac pacemakers			A27	Appropriate Device Problem Term/Code not Available		
1404	2007				CT257	Implantable cardiac pacemakers			A27	Appropriate Device Problem Term/Code not Available		
1405	2009				CT257	Implantable cardiac pacemakers			A27	Appropriate Device Problem Term/Code not Available		
1406	2008		47270	Cardiac resynchronization therapy implantable defibrillator	CT2866	Implantable cardiac electrical stimulators			A27	Appropriate Device Problem Term/Code not Available		
1407	2007				CT257	Implantable cardiac pacemakers			A27	Appropriate Device Problem Term/Code not Available		
1408	2008				CT257	Implantable cardiac pacemakers			A27	Appropriate Device Problem Term/Code not Available		
1409	2007				CT257	Implantable cardiac pacemakers			A27	Appropriate Device Problem Term/Code not Available		
1410	2008				CT257	Implantable cardiac pacemakers			A27	Appropriate Device Problem Term/Code not Available		
1411	2010				CT257	Implantable cardiac pacemakers			A26	Insufficient Device Problem Information		
1412	2010				CT257	Implantable cardiac pacemakers			A26	Insufficient Device Problem Information		
1413	2009		35223	Endocardial/interventricular septal pacing lead	CT1684	Cardiac pacemakers and associated devices			A27	Appropriate Device Problem Term/Code not Available		
1414	2006				CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A27	Appropriate Device Problem Term/Code not Available		
1415	2009				CT257	Implantable cardiac pacemakers			A27	Appropriate Device Problem Term/Code not Available		
1416	2006				CT257	Implantable cardiac pacemakers	F1905	Device Revision or Replacement	A27	Appropriate Device Problem Term/Code not Available		
1417	2008				CT257	Implantable cardiac pacemakers			A27	Appropriate Device Problem Term/Code not Available		
1418	2009				CT257	Implantable cardiac pacemakers			A27	Appropriate Device Problem Term/Code not Available		
1419	2008				CT257	Implantable cardiac pacemakers			A27	Appropriate Device Problem Term/Code not Available		
1420	2010				CT257	Implantable cardiac pacemakers			A27	Appropriate Device Problem Term/Code not Available		
1421	2008		47270	Cardiac resynchronization therapy implantable defibrillator	CT2866	Implantable cardiac electrical stimulators			A27	Appropriate Device Problem Term/Code not Available		
1422	2008				CT257	Implantable cardiac pacemakers			A27	Appropriate Device Problem Term/Code not Available		
1423	2008				CT257	Implantable cardiac pacemakers			A27	Appropriate Device Problem Term/Code not Available		
1424	2008		35223	Endocardial/interventricular septal pacing lead	CT1684	Cardiac pacemakers and associated devices			A27	Appropriate Device Problem Term/Code not Available		
1425	2008	All other reportable incidents			CT257	Implantable cardiac pacemakers			A27	Appropriate Device Problem Term/Code not Available		
1426	2008	All other reportable incidents			CT257	Implantable cardiac pacemakers			A27	Appropriate Device Problem Term/Code not Available		
1427	2008	All other reportable incidents			CT257	Implantable cardiac pacemakers			A27	Appropriate Device Problem Term/Code not Available		
1428	2006				CT257	Implantable cardiac pacemakers	F1905	Device Revision or Replacement	A27	Appropriate Device Problem Term/Code not Available		
1429	2009	All other reportable incidents			CT257	Implantable cardiac pacemakers			A27	Appropriate Device Problem Term/Code not Available		
1430	2010				CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A27	Appropriate Device Problem Term/Code not Available		
1431	2006				CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A27	Appropriate Device Problem Term/Code not Available		
1432	2009				CT257	Implantable cardiac pacemakers			A27	Appropriate Device Problem Term/Code not Available		
1433	2009		47270	Cardiac resynchronization therapy implantable defibrillator	CT2866	Implantable cardiac electrical stimulators			A27	Appropriate Device Problem Term/Code not Available		
1434	2007		35223	Endocardial/interventricular septal pacing lead	CT1684	Cardiac pacemakers and associated devices			A27	Appropriate Device Problem Term/Code not Available		
1435	2010				CT257	Implantable cardiac pacemakers			A27	Appropriate Device Problem Term/Code not Available		
1436	2009				CT257	Implantable cardiac pacemakers			A27	Appropriate Device Problem Term/Code not Available		
1437	2010				CT257	Implantable cardiac pacemakers			A26	Insufficient Device Problem Information		
1438	2008				CT257	Implantable cardiac pacemakers			A27	Appropriate Device Problem Term/Code not Available		
1439	2008				CT257	Implantable cardiac pacemakers			A27	Appropriate Device Problem Term/Code not Available		
1440	2008				CT257	Implantable cardiac pacemakers			A27	Appropriate Device Problem Term/Code not Available		
1441	2010				CT257	Implantable cardiac pacemakers			A27	Appropriate Device Problem Term/Code not Available		
1442	2010				CT257	Implantable cardiac pacemakers			A27	Appropriate Device Problem Term/Code not Available		
1443	2006				CT257	Implantable cardiac pacemakers			A27	Appropriate Device Problem Term/Code not Available		
1444	2006				CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A26	Insufficient Device Problem Information		
1445	2007				CT257	Implantable cardiac pacemakers			A27	Appropriate Device Problem Term/Code not Available		
1446	2007		47270	Cardiac resynchronization therapy implantable defibrillator	CT2866	Implantable cardiac electrical stimulators			A27	Appropriate Device Problem Term/Code not Available		
1447	2008				CT257	Implantable cardiac pacemakers			A27	Appropriate Device Problem Term/Code not Available		
1448	2009		47270	Cardiac resynchronization therapy implantable defibrillator	CT2866	Implantable cardiac electrical stimulators			A27	Appropriate Device Problem Term/Code not Available		
1449	2009				CT257	Implantable cardiac pacemakers			A26	Insufficient Device Problem Information		
1450	2009				CT257	Implantable cardiac pacemakers			A27	Appropriate Device Problem Term/Code not Available		
1451	2010				CT257	Implantable cardiac pacemakers			A26	Insufficient Device Problem Information		
1452	2006				CT257	Implantable cardiac pacemakers	F1905	Device Revision or Replacement	A27	Appropriate Device Problem Term/Code not Available		

1453	2008				CT257	Implantable cardiac pacemakers			A27	Appropriate Device Problem Term/Code not Available		
1454	2008				CT257	Implantable cardiac pacemakers			A27	Appropriate Device Problem Term/Code not Available		
1455	2006				CT257	Implantable cardiac pacemakers	F1905	Device Revision or Replacement	A27	Appropriate Device Problem Term/Code not Available		
1456	2006				CT257	Implantable cardiac pacemakers	F1905	Device Revision or Replacement	A27	Appropriate Device Problem Term/Code not Available		
1457	2007				CT257	Implantable cardiac pacemakers			A27	Appropriate Device Problem Term/Code not Available		
1458	2007				CT257	Implantable cardiac pacemakers			A27	Appropriate Device Problem Term/Code not Available		
1459	2006				CT257	Implantable cardiac pacemakers	F1905	Device Revision or Replacement	A27	Appropriate Device Problem Term/Code not Available		
1460	2006				CT257	Implantable cardiac pacemakers	F1905	Device Revision or Replacement	A27	Appropriate Device Problem Term/Code not Available		
1461	2008	All other reportable incidents			CT257	Implantable cardiac pacemakers			A27	Appropriate Device Problem Term/Code not Available		
1462	2006				CT257	Implantable cardiac pacemakers	F1905	Device Revision or Replacement	A27	Appropriate Device Problem Term/Code not Available		
1463	2010		47270	Cardiac resynchronization therapy implantable defibrillator	CT2866	Implantable cardiac electrical stimulators			A26	Insufficient Device Problem Information		
1464	2007				CT257	Implantable cardiac pacemakers			A27	Appropriate Device Problem Term/Code not Available		
1465	2009		35223	Endocardial/interventricular septal pacing lead	CT1684	Cardiac pacemakers and associated devices			A27	Appropriate Device Problem Term/Code not Available		
1466	2008				CT257	Implantable cardiac pacemakers			A27	Appropriate Device Problem Term/Code not Available		
1467	2008		35223	Endocardial/interventricular septal pacing lead	CT1684	Cardiac pacemakers and associated devices			A27	Appropriate Device Problem Term/Code not Available		
1468	2006		47270	Cardiac resynchronization therapy implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	F1905	Device Revision or Replacement	A27	Appropriate Device Problem Term/Code not Available		
1469	2009		47270	Cardiac resynchronization therapy implantable defibrillator	CT2866	Implantable cardiac electrical stimulators			A27	Appropriate Device Problem Term/Code not Available		
1470	2009				CT257	Implantable cardiac pacemakers			A27	Appropriate Device Problem Term/Code not Available		
1471	2009		35223	Endocardial/interventricular septal pacing lead	CT1684	Cardiac pacemakers and associated devices			A26	Insufficient Device Problem Information		
1472	2009	All other reportable incidents			CT257	Implantable cardiac pacemakers			A27	Appropriate Device Problem Term/Code not Available		
1473	2006				CT257	Implantable cardiac pacemakers	F1905	Device Revision or Replacement	A27	Appropriate Device Problem Term/Code not Available		
1474	2008				CT257	Implantable cardiac pacemakers			A27	Appropriate Device Problem Term/Code not Available		
1475	2006				CT257	Implantable cardiac pacemakers	F1905	Device Revision or Replacement	A27	Appropriate Device Problem Term/Code not Available		
1476	2008				CT257	Implantable cardiac pacemakers			A27	Appropriate Device Problem Term/Code not Available		
1477	2010				CT257	Implantable cardiac pacemakers			A26	Insufficient Device Problem Information		
1478	2009	All other reportable incidents			CT257	Implantable cardiac pacemakers			A27	Appropriate Device Problem Term/Code not Available		
1479	2009				CT257	Implantable cardiac pacemakers			A27	Appropriate Device Problem Term/Code not Available		
1480	2006				CT257	Implantable cardiac pacemakers	F1905	Device Revision or Replacement	A27	Appropriate Device Problem Term/Code not Available		
1481	2009		47270	Cardiac resynchronization therapy implantable defibrillator	CT2866	Implantable cardiac electrical stimulators			A27	Appropriate Device Problem Term/Code not Available		
1482	2009	All other reportable incidents			CT257	Implantable cardiac pacemakers			A26	Insufficient Device Problem Information	29/12/2008	
1483	2009				CT257	Implantable cardiac pacemakers			A27	Appropriate Device Problem Term/Code not Available		
1484	2006		47270	Cardiac resynchronization therapy implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	F1905	Device Revision or Replacement	A27	Appropriate Device Problem Term/Code not Available		
1485	2010				CT257	Implantable cardiac pacemakers			A26	Insufficient Device Problem Information		
1486	2010	All other reportable incidents			CT257	Implantable cardiac pacemakers			A27	Appropriate Device Problem Term/Code not Available		
1487	2009				CT257	Implantable cardiac pacemakers			A27	Appropriate Device Problem Term/Code not Available		
1488	2008				CT257	Implantable cardiac pacemakers			A27	Appropriate Device Problem Term/Code not Available		
1489	2009		35223	Endocardial/interventricular septal pacing lead	CT1684	Cardiac pacemakers and associated devices			A27	Appropriate Device Problem Term/Code not Available		
1490	2009				CT257	Implantable cardiac pacemakers			A27	Appropriate Device Problem Term/Code not Available		
1491	2008		35223	Endocardial/interventricular septal pacing lead	CT1684	Cardiac pacemakers and associated devices			A27	Appropriate Device Problem Term/Code not Available		
1492	2008		35223	Endocardial/interventricular septal pacing lead	CT1684	Cardiac pacemakers and associated devices			A27	Appropriate Device Problem Term/Code not Available		
1493	2008		47270	Cardiac resynchronization therapy implantable defibrillator	CT2866	Implantable cardiac electrical stimulators			A27	Appropriate Device Problem Term/Code not Available		
1494	2006				CT257	Implantable cardiac pacemakers	F1905	Device Revision or Replacement	A27	Appropriate Device Problem Term/Code not Available		
1495	2008		35223	Endocardial/interventricular septal pacing lead	CT1684	Cardiac pacemakers and associated devices			A27	Appropriate Device Problem Term/Code not Available		
1496	2008				CT257	Implantable cardiac pacemakers			A27	Appropriate Device Problem Term/Code not Available		
1497	2010				CT257	Implantable cardiac pacemakers			A07	Electrical /Electronic Property Problem		
1498	2008				CT257	Implantable cardiac pacemakers			A27	Appropriate Device Problem Term/Code not Available		
1499	2008		47270	Cardiac resynchronization therapy implantable defibrillator	CT2866	Implantable cardiac electrical stimulators			A27	Appropriate Device Problem Term/Code not Available		
1500	2009				CT257	Implantable cardiac pacemakers			A27	Appropriate Device Problem Term/Code not Available		
1501	2009		35223	Endocardial/interventricular septal pacing lead	CT1684	Cardiac pacemakers and associated devices			A27	Appropriate Device Problem Term/Code not Available		
1502	2010		47270	Cardiac resynchronization therapy implantable defibrillator	CT2866	Implantable cardiac electrical stimulators			A26	Insufficient Device Problem Information		
1503	2009		47270	Cardiac resynchronization therapy implantable defibrillator	CT2866	Implantable cardiac electrical stimulators			A27	Appropriate Device Problem Term/Code not Available		

1504	2010		47270	Cardiac resynchronization therapy implantable defibrillator	CT2866	Implantable cardiac electrical stimulators			A27	Appropriate Device Problem Term/Code not Available		
1505	2010				CT257	Implantable cardiac pacemakers			A27	Appropriate Device Problem Term/Code not Available		
1506	2007				CT257	Implantable cardiac pacemakers	F1905	Device Revision or Replacement	A27	Appropriate Device Problem Term/Code not Available		
1507	2009				CT257	Implantable cardiac pacemakers			A27	Appropriate Device Problem Term/Code not Available		
1508	2010		35223	Endocardial/interventricular septal pacing lead	CT1684	Cardiac pacemakers and associated devices			A27	Appropriate Device Problem Term/Code not Available		
1509	2009		47270	Cardiac resynchronization therapy implantable defibrillator	CT2866	Implantable cardiac electrical stimulators			A26	Insufficient Device Problem Information		
1510	2009		47270	Cardiac resynchronization therapy implantable defibrillator	CT2866	Implantable cardiac electrical stimulators			A26	Insufficient Device Problem Information		
1511	2008				CT257	Implantable cardiac pacemakers			A27	Appropriate Device Problem Term/Code not Available		
1512	2008		47270	Cardiac resynchronization therapy implantable defibrillator	CT2866	Implantable cardiac electrical stimulators			A27	Appropriate Device Problem Term/Code not Available		
1513	2010				CT257	Implantable cardiac pacemakers			A26	Insufficient Device Problem Information		
1514	2010		47270	Cardiac resynchronization therapy implantable defibrillator	CT2866	Implantable cardiac electrical stimulators			A27	Appropriate Device Problem Term/Code not Available		
1515	2010				CT257	Implantable cardiac pacemakers			A27	Appropriate Device Problem Term/Code not Available		
1516	2010		47270	Cardiac resynchronization therapy implantable defibrillator	CT2866	Implantable cardiac electrical stimulators			A27	Appropriate Device Problem Term/Code not Available		
1517	2009				CT257	Implantable cardiac pacemakers			A27	Appropriate Device Problem Term/Code not Available		
1518	2008				CT257	Implantable cardiac pacemakers			A27	Appropriate Device Problem Term/Code not Available		
1519	2010		47270	Cardiac resynchronization therapy implantable defibrillator	CT2866	Implantable cardiac electrical stimulators			A27	Appropriate Device Problem Term/Code not Available		
1520	2007				CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A27	Appropriate Device Problem Term/Code not Available		
1521	2009				CT257	Implantable cardiac pacemakers			A27	Appropriate Device Problem Term/Code not Available		
1522	2009		35223	Endocardial/interventricular septal pacing lead	CT1684	Cardiac pacemakers and associated devices			A27	Appropriate Device Problem Term/Code not Available		
1523	2008				CT257	Implantable cardiac pacemakers			A27	Appropriate Device Problem Term/Code not Available		
1524	2008				CT257	Implantable cardiac pacemakers			A27	Appropriate Device Problem Term/Code not Available		
1525	2007				CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A26	Insufficient Device Problem Information		
1526	2009		47270	Cardiac resynchronization therapy implantable defibrillator	CT2866	Implantable cardiac electrical stimulators			A26	Insufficient Device Problem Information		
1527	2007		47270	Cardiac resynchronization therapy implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	F1905	Device Revision or Replacement	A27	Appropriate Device Problem Term/Code not Available		
1528	2009				CT257	Implantable cardiac pacemakers			A27	Appropriate Device Problem Term/Code not Available		
1529	2009				CT257	Implantable cardiac pacemakers			A27	Appropriate Device Problem Term/Code not Available		
1530	2009				CT257	Implantable cardiac pacemakers			A27	Appropriate Device Problem Term/Code not Available		
1531	2009	Unanticipated serious deterioration in state of health			CT257	Implantable cardiac pacemakers			A27	Appropriate Device Problem Term/Code not Available		
1532	2009	All other reportable incidents			CT257	Implantable cardiac pacemakers			A27	Appropriate Device Problem Term/Code not Available		
1533	2008				CT257	Implantable cardiac pacemakers			A27	Appropriate Device Problem Term/Code not Available		
1534	2008				CT257	Implantable cardiac pacemakers			A27	Appropriate Device Problem Term/Code not Available		
1535	2010				CT257	Implantable cardiac pacemakers			A27	Appropriate Device Problem Term/Code not Available		
1536	2010		47270	Cardiac resynchronization therapy implantable defibrillator	CT2866	Implantable cardiac electrical stimulators			A27	Appropriate Device Problem Term/Code not Available		
1537	2008				CT257	Implantable cardiac pacemakers			A27	Appropriate Device Problem Term/Code not Available		
1538	2009		47270	Cardiac resynchronization therapy implantable defibrillator	CT2866	Implantable cardiac electrical stimulators			A27	Appropriate Device Problem Term/Code not Available		
1539	2008				CT257	Implantable cardiac pacemakers			A27	Appropriate Device Problem Term/Code not Available		
1540	2008				CT257	Implantable cardiac pacemakers			A27	Appropriate Device Problem Term/Code not Available		
1541	2008				CT257	Implantable cardiac pacemakers			A27	Appropriate Device Problem Term/Code not Available		
1542	2008		47270	Cardiac resynchronization therapy implantable defibrillator	CT2866	Implantable cardiac electrical stimulators			A27	Appropriate Device Problem Term/Code not Available		
1543	2008				CT257	Implantable cardiac pacemakers			A27	Appropriate Device Problem Term/Code not Available		
1544	2010				CT257	Implantable cardiac pacemakers			A27	Appropriate Device Problem Term/Code not Available		
1545	2009		47270	Cardiac resynchronization therapy implantable defibrillator	CT2866	Implantable cardiac electrical stimulators			A27	Appropriate Device Problem Term/Code not Available		
1546	2008		47270	Cardiac resynchronization therapy implantable defibrillator	CT2866	Implantable cardiac electrical stimulators			A27	Appropriate Device Problem Term/Code not Available		
1547	2007				CT257	Implantable cardiac pacemakers	F1905	Device Revision or Replacement	A27	Appropriate Device Problem Term/Code not Available		
1548	2007				CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A27	Appropriate Device Problem Term/Code not Available		
1549	2009		35223	Endocardial/interventricular septal pacing lead	CT1684	Cardiac pacemakers and associated devices			A27	Appropriate Device Problem Term/Code not Available		
1550	2009				CT257	Implantable cardiac pacemakers			A27	Appropriate Device Problem Term/Code not Available		
1551	2010	All other reportable incidents			CT257	Implantable cardiac pacemakers			A27	Appropriate Device Problem Term/Code not Available		
1552	2010	All other reportable incidents			CT257	Implantable cardiac pacemakers			A27	Appropriate Device Problem Term/Code not Available		
1553	2007				CT257	Implantable cardiac pacemakers	F1905	Device Revision or Replacement	A27	Appropriate Device Problem Term/Code not Available		

1554	2010				CT257	Implantable cardiac pacemakers			A27	Appropriate Device Problem Term/Code not Available		
1555	2010		35223	Endocardial/interventricular septal pacing lead	CT1684	Cardiac pacemakers and associated devices			A26	Insufficient Device Problem Information		
1556	2009				CT257	Implantable cardiac pacemakers			A27	Appropriate Device Problem Term/Code not Available		
1557	2007				CT257	Implantable cardiac pacemakers	F1905	Device Revision or Replacement	A27	Appropriate Device Problem Term/Code not Available		
1558	2010				CT257	Implantable cardiac pacemakers			A27	Appropriate Device Problem Term/Code not Available		
1559	2009	Unanticipated serious deterioration in state of health			CT257	Implantable cardiac pacemakers			A27	Appropriate Device Problem Term/Code not Available		
1560	2010				CT257	Implantable cardiac pacemakers			A0104	Migration or Expulsion of Device		
1561	2010		47270	Cardiac resynchronization therapy implantable defibrillator	CT2866	Implantable cardiac electrical stimulators			A26	Insufficient Device Problem Information		
1562	2010				CT257	Implantable cardiac pacemakers			A26	Insufficient Device Problem Information		
1563	2010				CT257	Implantable cardiac pacemakers			A09	Output Problem		
									A26	Insufficient Device Problem Information		
1564	2007				CT257	Implantable cardiac pacemakers	F1905	Device Revision or Replacement	A27	Appropriate Device Problem Term/Code not Available		
1565	2010				CT257	Implantable cardiac pacemakers			A11	Computer Software Problem		
									A26	Insufficient Device Problem Information		
1566	2010				CT257	Implantable cardiac pacemakers			A12	Connection Problem		
1567	2010				CT257	Implantable cardiac pacemakers			A26	Insufficient Device Problem Information		
1568	2010				CT257	Implantable cardiac pacemakers			A26	Insufficient Device Problem Information		
1569	2010				CT257	Implantable cardiac pacemakers			A07	Electrical /Electronic Property Problem		
									A09	Output Problem		
1570	2010				CT257	Implantable cardiac pacemakers			A26	Insufficient Device Problem Information		
1571	2010				CT257	Implantable cardiac pacemakers			A04	Material Integrity Problem		
1572	2009	All other reportable incidents			CT257	Implantable cardiac pacemakers			A27	Appropriate Device Problem Term/Code not Available		
1573	2010				CT257	Implantable cardiac pacemakers			A27	Appropriate Device Problem Term/Code not Available		
1574	2010		47270	Cardiac resynchronization therapy implantable defibrillator	CT2866	Implantable cardiac electrical stimulators			A27	Appropriate Device Problem Term/Code not Available		
1575	2010				CT257	Implantable cardiac pacemakers	F1905	Device Revision or Replacement	A11	Computer Software Problem		
									A26	Insufficient Device Problem Information		
1576	2010				CT257	Implantable cardiac pacemakers			A26	Insufficient Device Problem Information		
1577	2010		47270	Cardiac resynchronization therapy implantable defibrillator	CT2866	Implantable cardiac electrical stimulators			A27	Appropriate Device Problem Term/Code not Available		
1578	2010		47270	Cardiac resynchronization therapy implantable defibrillator	CT2866	Implantable cardiac electrical stimulators			A04	Material Integrity Problem		
1579	2010		47270	Cardiac resynchronization therapy implantable defibrillator	CT2866	Implantable cardiac electrical stimulators			A26	Insufficient Device Problem Information		
1580	2010		35223	Endocardial/interventricular septal pacing lead	CT1684	Cardiac pacemakers and associated devices			A26	Insufficient Device Problem Information		
1581	2010				CT257	Implantable cardiac pacemakers			A27	Appropriate Device Problem Term/Code not Available		
1582	2010				CT257	Implantable cardiac pacemakers			A0104	Migration or Expulsion of Device		
1583	2010	Death			CT257	Implantable cardiac pacemakers			A05	Mechanical Problem	19/11/2010	
1584	2010				CT257	Implantable cardiac pacemakers			A27	Appropriate Device Problem Term/Code not Available		
1585	2010				CT257	Implantable cardiac pacemakers			A26	Insufficient Device Problem Information		
1586	2010				CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A05	Mechanical Problem		
1587	2010				CT257	Implantable cardiac pacemakers			A27	Appropriate Device Problem Term/Code not Available		
1588	2010				CT257	Implantable cardiac pacemakers			A27	Appropriate Device Problem Term/Code not Available		
1589	2010				CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A04	Material Integrity Problem		
1590	2010				CT257	Implantable cardiac pacemakers	F1905	Device Revision or Replacement	A23	Use of Device Problem		
1591	2010				CT257	Implantable cardiac pacemakers			A27	Appropriate Device Problem Term/Code not Available		
1592	2010				CT257	Implantable cardiac pacemakers			A27	Appropriate Device Problem Term/Code not Available		
1593	2010		47270	Cardiac resynchronization therapy implantable defibrillator	CT2866	Implantable cardiac electrical stimulators			A27	Appropriate Device Problem Term/Code not Available		
1594	2010		47270	Cardiac resynchronization therapy implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	F1905	Device Revision or Replacement	A14	Infusion or Flow Problem		
1595	2010		35223	Endocardial/interventricular septal pacing lead	CT1684	Cardiac pacemakers and associated devices			A26	Insufficient Device Problem Information		
1596	2010				CT257	Implantable cardiac pacemakers			A26	Insufficient Device Problem Information		
1597	2010				CT257	Implantable cardiac pacemakers			A14	Infusion or Flow Problem		
1598	2010		35223	Endocardial/interventricular septal pacing lead	CT1684	Cardiac pacemakers and associated devices			A14	Infusion or Flow Problem		
1599	2010	All other reportable incidents			CT257	Implantable cardiac pacemakers			A27	Appropriate Device Problem Term/Code not Available		
1600	2010		47270	Cardiac resynchronization therapy implantable defibrillator	CT2866	Implantable cardiac electrical stimulators			A27	Appropriate Device Problem Term/Code not Available		
1601	2010				CT257	Implantable cardiac pacemakers			A27	Appropriate Device Problem Term/Code not Available		
1602	2010				CT257	Implantable cardiac pacemakers			A27	Appropriate Device Problem Term/Code not Available		
1603	2010				CT257	Implantable cardiac pacemakers			A04	Material Integrity Problem		
1604	2010	All other reportable incidents			CT257	Implantable cardiac pacemakers			A04	Material Integrity Problem		
1605	2010		47270	Cardiac resynchronization therapy implantable defibrillator	CT2866	Implantable cardiac electrical stimulators			A26	Insufficient Device Problem Information		
1606	2011		47270	Cardiac resynchronization therapy implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	F1905	Device Revision or Replacement	A04	Material Integrity Problem		
									A0404	Crack		
									A16	Protective Measures Problem		
									A1602	Fail-Safe Problem		
1607	2011		47270	Cardiac resynchronization therapy implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E2401	Insufficient Information	A07	Electrical /Electronic Property Problem		
									A0709	Device Sensing Problem		
									A11	Computer Software Problem		
									A1102	Application Program Problem		
1608	2011		47270	Cardiac resynchronization therapy implantable	CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A07	Electrical /Electronic Property Problem		

				defibrillator					A0721	Circuit Failure		
1609	2011				CT257	Implantable cardiac pacemakers			A09	Output Problem		
									A0907	No Device Output		
1610	2011		47270	Cardiac resynchronization therapy implantable defibrillator	CT2866	Implantable cardiac electrical stimulators			A26	Insufficient Device Problem Information		
1611	2012				CT257	Implantable cardiac pacemakers	F15	Recognized Device or Procedural Complication	A26	Insufficient Device Problem Information		
1612	2012				CT257	Implantable cardiac pacemakers	F1905	Device Revision or Replacement	A13	Communication or Transmission Problem		
									A26	Insufficient Device Problem Information		
1613	2012		47270	Cardiac resynchronization therapy implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	F15	Recognized Device or Procedural Complication	A26	Insufficient Device Problem Information		
1614	2011	All other reportable incidents			CT257	Implantable cardiac pacemakers	F1905	Device Revision or Replacement	A1304	Telemetry Discrepancy	02/06/2011	
									A26	Insufficient Device Problem Information	02/06/2011	
1615	2012				CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A26	Insufficient Device Problem Information		
1616	2011	All other reportable incidents			CT257	Implantable cardiac pacemakers	F1905	Device Revision or Replacement	A07	Electrical /Electronic Property Problem	07/01/2011	
									A0708	Power Problem	07/01/2011	
1617	2012	All other reportable incidents			CT257	Implantable cardiac pacemakers			A07	Electrical /Electronic Property Problem		
									A0708	Power Problem		
1618	2011		35039	Epicardial pacing lead	CT1684	Cardiac pacemakers and associated devices			A07	Electrical /Electronic Property Problem		
									A0709	Device Sensing Problem		
1619	2012		47270	Cardiac resynchronization therapy implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A07	Electrical /Electronic Property Problem		
									A0709	Device Sensing Problem		
1620	2011				CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A07	Electrical /Electronic Property Problem		
									A0709	Device Sensing Problem		
1621	2012		47270	Cardiac resynchronization therapy implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	F15	Recognized Device or Procedural Complication	A26	Insufficient Device Problem Information		
1622	2012		47270	Cardiac resynchronization therapy implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E2401	Insufficient Information	A26	Insufficient Device Problem Information		
1623	2012		47270	Cardiac resynchronization therapy implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E2401	Insufficient Information	A26	Insufficient Device Problem Information		
1624	2012		47270	Cardiac resynchronization therapy implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E2401	Insufficient Information	A26	Insufficient Device Problem Information		
1625	2011	All other reportable incidents			CT257	Implantable cardiac pacemakers			A07	Electrical /Electronic Property Problem	07/01/2011	
									A0708	Power Problem	07/01/2011	
1626	2012		47270	Cardiac resynchronization therapy implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E2401	Insufficient Information	A26	Insufficient Device Problem Information		
1627	2012		47270	Cardiac resynchronization therapy implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E2401	Insufficient Information	A26	Insufficient Device Problem Information		
1628	2012		47270	Cardiac resynchronization therapy implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E2401	Insufficient Information	A26	Insufficient Device Problem Information		
1629	2012		47270	Cardiac resynchronization therapy implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E2401	Insufficient Information	A26	Insufficient Device Problem Information		
1630	2012		47270	Cardiac resynchronization therapy implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	F21	Unexpected Deterioration	A26	Insufficient Device Problem Information		
1631	2012		47270	Cardiac resynchronization therapy implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	F21	Unexpected Deterioration	A26	Insufficient Device Problem Information		
1632	2012		47270	Cardiac resynchronization therapy implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	F21	Unexpected Deterioration	A26	Insufficient Device Problem Information		
1633	2012		35223	Endocardial/interventricular septal pacing lead	CT1684	Cardiac pacemakers and associated devices	F1905	Device Revision or Replacement	A15	Activation, Positioning or Separation Problem		
									A150203	Difficult or Delayed Positioning		
1634	2012				CT257	Implantable cardiac pacemakers	E0514	Thrombosis/Thrombus	A07	Electrical /Electronic Property Problem		
									A0709	Device Sensing Problem		
							E1906	Unspecified Infection	A07	Electrical /Electronic Property Problem		
									A0709	Device Sensing Problem		
							E2401	Insufficient Information	A07	Electrical /Electronic Property Problem		
									A0709	Device Sensing Problem		
1635	2011	All other reportable incidents			CT257	Implantable cardiac pacemakers	F1905	Device Revision or Replacement	A07	Electrical /Electronic Property Problem	03/06/2011	
									A0708	Power Problem	03/06/2011	
1636	2011	All other reportable incidents			CT257	Implantable cardiac pacemakers	F1905	Device Revision or Replacement	A09	Output Problem	28/04/2011	
									A0907	No Device Output	28/04/2011	
									A1304	Telemetry Discrepancy	28/04/2011	
									A26	Insufficient Device Problem Information	28/04/2011	
1637	2011		47270	Cardiac resynchronization therapy implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E2401	Insufficient Information	A26	Insufficient Device Problem Information		
1638	2011		47270	Cardiac resynchronization therapy implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E2401	Insufficient Information	A26	Insufficient Device Problem Information		
1639	2011		47270	Cardiac resynchronization therapy implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E2401	Insufficient Information	A26	Insufficient Device Problem Information		
1640	2011		47270	Cardiac resynchronization therapy implantable defibrillator	CT2866	Implantable cardiac electrical stimulators			A26	Insufficient Device Problem Information		
1641	2011		47270	Cardiac resynchronization therapy implantable defibrillator	CT2866	Implantable cardiac electrical stimulators			A26	Insufficient Device Problem Information		
1642	2012		35223	Endocardial/interventricular septal pacing lead	CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A04	Material Integrity Problem		
									A0405	Degraded		
1643	2012		47270	Cardiac resynchronization therapy implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	F1905	Device Revision or Replacement	A09	Output Problem		
									A090404	Output Above Specifications		
1644	2011		47270	Cardiac resynchronization therapy implantable defibrillator	CT2866	Implantable cardiac electrical stimulators			A26	Insufficient Device Problem Information		
1645	2011		47270	Cardiac resynchronization therapy implantable defibrillator	CT2866	Implantable cardiac electrical stimulators			A27	Appropriate Device Problem Term/Code not Available		
1646	2011		47270	Cardiac resynchronization therapy implantable defibrillator	CT2866	Implantable cardiac electrical stimulators			A26	Insufficient Device Problem Information		
1647	2012				CT257	Implantable cardiac pacemakers	F1905	Device Revision or Replacement	A07	Electrical /Electronic Property Problem		
									A0708	Power Problem		
1648	2012				CT257	Implantable cardiac pacemakers	F1905	Device Revision or Replacement	A07	Electrical /Electronic Property Problem		
									A0708	Power Problem		
									A13	Communication or Transmission Problem		
									A26	Insufficient Device Problem Information		
1649	2011				CT257	Implantable cardiac pacemakers			A27	Appropriate Device Problem Term/Code not Available		
1650	2012				CT257	Implantable cardiac pacemakers	E2401	Insufficient Information	A26	Insufficient Device Problem Information		
1651	2012				CT257	Implantable cardiac pacemakers	F1905	Device Revision or Replacement	A07	Electrical /Electronic Property Problem		
									A0708	Power Problem		

1652	2012		47270	Cardiac resynchronization therapy implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E2401	Insufficient Information	A26	Insufficient Device Problem Information		
1653	2011				CT257	Implantable cardiac pacemakers			A07	Electrical /Electronic Property Problem	11/01/2011	
									A0708	Power Problem	11/01/2011	
1654	2012				CT257	Implantable cardiac pacemakers	F15	Recognized Device or Procedural Complication	A26	Insufficient Device Problem Information		
1655	2011		47270	Cardiac resynchronization therapy implantable defibrillator	CT2866	Implantable cardiac electrical stimulators			A26	Insufficient Device Problem Information		
1656	2012		47270	Cardiac resynchronization therapy implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E2401	Insufficient Information	A26	Insufficient Device Problem Information		
1657	2012		47270	Cardiac resynchronization therapy implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E2401	Insufficient Information	A26	Insufficient Device Problem Information		
1658	2012		47270	Cardiac resynchronization therapy implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E2401	Insufficient Information	A26	Insufficient Device Problem Information		
1659	2012		47270	Cardiac resynchronization therapy implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E2401	Insufficient Information	A26	Insufficient Device Problem Information		
1660	2012		47270	Cardiac resynchronization therapy implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E2401	Insufficient Information	A26	Insufficient Device Problem Information		
1661	2011				CT257	Implantable cardiac pacemakers	F1905	Device Revision or Replacement	A27	Appropriate Device Problem Term/Code not Available		
1662	2011				CT257	Implantable cardiac pacemakers	F1905	Device Revision or Replacement	A07	Electrical /Electronic Property Problem		
									A0708	Power Problem		
1663	2011				CT257	Implantable cardiac pacemakers			A27	Appropriate Device Problem Term/Code not Available		
1664	2011				CT257	Implantable cardiac pacemakers	F1905	Device Revision or Replacement	A09	Output Problem		
									A0908	Incorrect, Inadequate or Imprecise Result or Readings		
1665	2011				CT257	Implantable cardiac pacemakers	E2401	Insufficient Information	A07	Electrical /Electronic Property Problem		
									A0708	Power Problem		
1666	2011		47270	Cardiac resynchronization therapy implantable defibrillator	CT2866	Implantable cardiac electrical stimulators			A26	Insufficient Device Problem Information		
1667	2011				CT257	Implantable cardiac pacemakers	F1905	Device Revision or Replacement	A07	Electrical /Electronic Property Problem		
									A0708	Power Problem		
1668	2011				CT257	Implantable cardiac pacemakers	F1905	Device Revision or Replacement	A07	Electrical /Electronic Property Problem		
									A0721	Circuit Failure		
1669	2012		47270	Cardiac resynchronization therapy implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	F1905	Device Revision or Replacement	A07	Electrical /Electronic Property Problem		
									A0708	Power Problem		
1670	2011	All other reportable incidents			CT257	Implantable cardiac pacemakers	F1905	Device Revision or Replacement	A07	Electrical /Electronic Property Problem		
									A0708	Power Problem		
1671	2011		47270	Cardiac resynchronization therapy implantable defibrillator	CT2866	Implantable cardiac electrical stimulators			A13	Communication or Transmission Problem		
									A26	Insufficient Device Problem Information		
1672	2011				CT257	Implantable cardiac pacemakers	F1905	Device Revision or Replacement	A09	Output Problem		
									A0907	No Device Output		
1673	2012		47270	Cardiac resynchronization therapy implantable defibrillator	CT2866	Implantable cardiac electrical stimulators			A26	Insufficient Device Problem Information		
1674	2012		47270	Cardiac resynchronization therapy implantable defibrillator	CT2866	Implantable cardiac electrical stimulators			A26	Insufficient Device Problem Information		
1675	2012		47270	Cardiac resynchronization therapy implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E2401	Insufficient Information	A26	Insufficient Device Problem Information		
1676	2011		47270	Cardiac resynchronization therapy implantable defibrillator	CT2866	Implantable cardiac electrical stimulators			A26	Insufficient Device Problem Information		
1677	2011		35223	Endocardial/interventricular septal pacing lead	CT1684	Cardiac pacemakers and associated devices			A09	Output Problem		
									A0907	No Device Output		
1678	2011		47270	Cardiac resynchronization therapy implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A26	Insufficient Device Problem Information		
1679	2011				CT257	Implantable cardiac pacemakers			A27	Appropriate Device Problem Term/Code not Available		
1680	2011		47270	Cardiac resynchronization therapy implantable defibrillator	CT2866	Implantable cardiac electrical stimulators			A26	Insufficient Device Problem Information		
1681	2012		47270	Cardiac resynchronization therapy implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	F15	Recognized Device or Procedural Complication	A26	Insufficient Device Problem Information		
1682	2011	All other reportable incidents			CT257	Implantable cardiac pacemakers	F1905	Device Revision or Replacement	A09	Output Problem		
									A0907	No Device Output		
1683	2011		47270	Cardiac resynchronization therapy implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	F1905	Device Revision or Replacement	A0104	Migration or Expulsion of Device		
									A07	Electrical /Electronic Property Problem		
									A0709	Device Sensing Problem		
1684	2011		47270	Cardiac resynchronization therapy implantable defibrillator	CT2866	Implantable cardiac electrical stimulators			A26	Insufficient Device Problem Information		
1685	2012		47270	Cardiac resynchronization therapy implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E2401	Insufficient Information	A26	Insufficient Device Problem Information		
1686	2012		47270	Cardiac resynchronization therapy implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E2401	Insufficient Information	A26	Insufficient Device Problem Information		
1687	2012		47270	Cardiac resynchronization therapy implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E2401	Insufficient Information	A26	Insufficient Device Problem Information		
1688	2012		47270	Cardiac resynchronization therapy implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E0402	Hypersensitivity/Allergic Reaction	A26	Insufficient Device Problem Information		
1689	2012		47270	Cardiac resynchronization therapy implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E2401	Insufficient Information	A26	Insufficient Device Problem Information		
1690	2012		47270	Cardiac resynchronization therapy implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E2401	Insufficient Information	A26	Insufficient Device Problem Information		
1691	2012		47270	Cardiac resynchronization therapy implantable defibrillator	CT2866	Implantable cardiac electrical stimulators			A26	Insufficient Device Problem Information		
1692	2012		47270	Cardiac resynchronization therapy implantable defibrillator	CT2866	Implantable cardiac electrical stimulators			A26	Insufficient Device Problem Information		
1693	2012		47270	Cardiac resynchronization therapy implantable defibrillator	CT2866	Implantable cardiac electrical stimulators			A26	Insufficient Device Problem Information		
1694	2012		47270	Cardiac resynchronization therapy implantable defibrillator	CT2866	Implantable cardiac electrical stimulators			A26	Insufficient Device Problem Information		
1695	2011				CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A27	Appropriate Device Problem Term/Code not Available		
1696	2011	All other reportable incidents			CT257	Implantable cardiac pacemakers	F1905	Device Revision or Replacement	A07	Electrical /Electronic Property Problem	09/11/2011	
									A0708	Power Problem	09/11/2011	
1697	2011				CT257	Implantable cardiac pacemakers	F1905	Device Revision or Replacement	A13	Communication or Transmission Problem		
									A26	Insufficient Device Problem Information		
1698	2012				CT257	Implantable cardiac pacemakers	E0514	Thrombosis/Thrombus	A07	Electrical /Electronic Property Problem		
									A0709	Device Sensing Problem		

									A26	Insufficient Device Problem Information		
							E1906	Unspecified Infection	A07	Electrical /Electronic Property Problem		
									A0709	Device Sensing Problem		
									A26	Insufficient Device Problem Information		
							E2401	Insufficient Information	A07	Electrical /Electronic Property Problem		
									A0709	Device Sensing Problem		
									A26	Insufficient Device Problem Information		
1699	2011		47270	Cardiac resynchronization therapy implantable defibrillator	CT2866	Implantable cardiac electrical stimulators			A26	Insufficient Device Problem Information		
1700	2011		35223	Endocardial/interventricular septal pacing lead	CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A15	Activation, Positioning or Separation Problem		
									A150301	Separation Failure		
1701	2011				CT257	Implantable cardiac pacemakers			A27	Appropriate Device Problem Term/Code not Available		
1702	2011		47270	Cardiac resynchronization therapy implantable defibrillator	CT2866	Implantable cardiac electrical stimulators			A27	Appropriate Device Problem Term/Code not Available		
1703	2011				CT257	Implantable cardiac pacemakers	E2401	Insufficient Information	A1304	Telemetry Discrepancy		
									A26	Insufficient Device Problem Information		
1704	2012				CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A07	Electrical /Electronic Property Problem		
									A0709	Device Sensing Problem		
1705	2012				CT257	Implantable cardiac pacemakers			A09	Output Problem		
									A090404	Output Above Specifications		
1706	2011				CT257	Implantable cardiac pacemakers	F1905	Device Revision or Replacement	A1304	Telemetry Discrepancy		
									A26	Insufficient Device Problem Information		
1707	2011				CT257	Implantable cardiac pacemakers	F1905	Device Revision or Replacement	A07	Electrical /Electronic Property Problem		
									A0709	Device Sensing Problem		
1708	2011				CT257	Implantable cardiac pacemakers			A07	Electrical /Electronic Property Problem		
									A0708	Power Problem		
1709	2011				CT257	Implantable cardiac pacemakers			A11	Computer Software Problem		
									A1102	Application Program Problem		
									A13	Communication or Transmission Problem		
									A26	Insufficient Device Problem Information		
1710	2011		47270	Cardiac resynchronization therapy implantable defibrillator	CT2866	Implantable cardiac electrical stimulators			A26	Insufficient Device Problem Information		
1711	2011		47270	Cardiac resynchronization therapy implantable defibrillator	CT2866	Implantable cardiac electrical stimulators			A26	Insufficient Device Problem Information		
1712	2011	All other reportable incidents			CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A09	Output Problem		
									A0907	No Device Output		
									A1304	Telemetry Discrepancy		
									A26	Insufficient Device Problem Information		
1713	2012		47270	Cardiac resynchronization therapy implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	F1905	Device Revision or Replacement	A0104	Migration or Expulsion of Device		
1714	2012				CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A07	Electrical /Electronic Property Problem		
									A0708	Power Problem		
1715	2011				CT1684	Cardiac pacemakers and associated devices			A07	Electrical /Electronic Property Problem		
									A0708	Power Problem		
1716	2012				CT257	Implantable cardiac pacemakers	F1905	Device Revision or Replacement	A07	Electrical /Electronic Property Problem		
									A0708	Power Problem		
1717	2012				CT257	Implantable cardiac pacemakers	E2401	Insufficient Information	A07	Electrical /Electronic Property Problem		
									A0709	Device Sensing Problem		
									A09	Output Problem		
									A0907	No Device Output		
1718	2011				CT257	Implantable cardiac pacemakers			A26	Insufficient Device Problem Information		
1719	2011		47270	Cardiac resynchronization therapy implantable defibrillator	CT2866	Implantable cardiac electrical stimulators			A26	Insufficient Device Problem Information		
1720	2011		47270	Cardiac resynchronization therapy implantable defibrillator	CT2866	Implantable cardiac electrical stimulators			A26	Insufficient Device Problem Information		
1721	2011		35223	Endocardial/interventricular septal pacing lead	CT1684	Cardiac pacemakers and associated devices			A09	Output Problem		
									A0907	No Device Output		
1722	2011				CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A27	Appropriate Device Problem Term/Code not Available		
1723	2012				CT257	Implantable cardiac pacemakers	F1905	Device Revision or Replacement	A07	Electrical /Electronic Property Problem		
									A0708	Power Problem		
1724	2012		47270	Cardiac resynchronization therapy implantable defibrillator	CT2866	Implantable cardiac electrical stimulators			A26	Insufficient Device Problem Information		
1725	2012		47270	Cardiac resynchronization therapy implantable defibrillator	CT2866	Implantable cardiac electrical stimulators			A26	Insufficient Device Problem Information		
1726	2012				CT257	Implantable cardiac pacemakers			A0104	Migration or Expulsion of Device		
1727	2011		47270	Cardiac resynchronization therapy implantable defibrillator	CT2866	Implantable cardiac electrical stimulators			A27	Appropriate Device Problem Term/Code not Available		
1728	2011		47270	Cardiac resynchronization therapy implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E2401	Insufficient Information	A04	Material Integrity Problem		
									A0405	Degraded		
1729	2012		47263	Cardiac resynchronization therapy implantable pacemaker	CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A07	Electrical /Electronic Property Problem		
									A0709	Device Sensing Problem		
					CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A07	Electrical /Electronic Property Problem		
									A0709	Device Sensing Problem		
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A07	Electrical /Electronic Property Problem		
									A0709	Device Sensing Problem		
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A07	Electrical /Electronic Property Problem		
									A0709	Device Sensing Problem		
1730	2012				CT257	Implantable cardiac pacemakers	F1905	Device Revision or Replacement	A07	Electrical /Electronic Property Problem		
									A0721	Circuit Failure		
1731	2012	All other reportable incidents	35223	Endocardial/interventricular septal pacing lead	CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A07	Electrical /Electronic Property Problem		
									A0709	Device Sensing Problem		
1732	2012		47270	Cardiac resynchronization therapy implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	F15	Recognized Device or Procedural Complication	A26	Insufficient Device Problem Information		
1733	2012		47270	Cardiac resynchronization therapy implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	F15	Recognized Device or Procedural Complication	A26	Insufficient Device Problem Information		
1734	2012		47270	Cardiac resynchronization therapy implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A26	Insufficient Device Problem Information		
1735	2012		47270	Cardiac resynchronization therapy implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	F21	Unexpected Deterioration	A26	Insufficient Device Problem Information		

1736	2012		47270	Cardiac resynchronization therapy implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E2401	Insufficient Information	A26	Insufficient Device Problem Information		
1737	2012		47270	Cardiac resynchronization therapy implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E2401	Insufficient Information	A26	Insufficient Device Problem Information		
1738	2012		47270	Cardiac resynchronization therapy implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	F21	Unexpected Deterioration	A26	Insufficient Device Problem Information		
1739	2012		47270	Cardiac resynchronization therapy implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E0402	Hypersensitivity/Allergic Reaction	A26	Insufficient Device Problem Information		
1740	2012		47270	Cardiac resynchronization therapy implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A26	Insufficient Device Problem Information		
1741	2012		47270	Cardiac resynchronization therapy implantable defibrillator	CT2866	Implantable cardiac electrical stimulators			A26	Insufficient Device Problem Information		
1742	2012		47270	Cardiac resynchronization therapy implantable defibrillator	CT2866	Implantable cardiac electrical stimulators			A26	Insufficient Device Problem Information		
1743	2012		47270	Cardiac resynchronization therapy implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	F21	Unexpected Deterioration	A26	Insufficient Device Problem Information		
1744	2012		47270	Cardiac resynchronization therapy implantable defibrillator	CT2866	Implantable cardiac electrical stimulators			A26	Insufficient Device Problem Information		
1745	2011				CT257	Implantable cardiac pacemakers	E2401	Insufficient Information	A07	Electrical /Electronic Property Problem		
									A0721	Circuit Failure		
1746	2011		47270	Cardiac resynchronization therapy implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	F1905	Device Revision or Replacement	A09	Output Problem		
									A0907	No Device Output		
1747	2012		47270	Cardiac resynchronization therapy implantable defibrillator	CT2866	Implantable cardiac electrical stimulators			A26	Insufficient Device Problem Information		
1748	2012		47270	Cardiac resynchronization therapy implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	F15	Recognized Device or Procedural Complication	A26	Insufficient Device Problem Information		
1749	2012	All other reportable incidents			CT257	Implantable cardiac pacemakers	F1905	Device Revision or Replacement	A07	Electrical /Electronic Property Problem	29/07/2012	
									A0708	Power Problem	29/07/2012	
1750	2011	All other reportable incidents			CT257	Implantable cardiac pacemakers			A09	Output Problem	29/06/2011	
									A0907	No Device Output	29/06/2011	
1751	2011		47270	Cardiac resynchronization therapy implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E2401	Insufficient Information	A12	Connection Problem		
1752	2012				CT257	Implantable cardiac pacemakers	E2401	Insufficient Information	A09	Output Problem		
									A0907	No Device Output		
1753	2012		35223	Endocardial/interventricular septal pacing lead	CT1684	Cardiac pacemakers and associated devices	F1905	Device Revision or Replacement	A04	Material Integrity Problem		
									A0405	Degraded		
1754	2011		35223	Endocardial/interventricular septal pacing lead	CT1684	Cardiac pacemakers and associated devices			A26	Insufficient Device Problem Information		
1755	2011		47270	Cardiac resynchronization therapy implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	F21	Unexpected Deterioration	A26	Insufficient Device Problem Information		
1756	2012				CT257	Implantable cardiac pacemakers	F1905	Device Revision or Replacement	A04	Material Integrity Problem		
									A0413	Material Separation		
1757	2012		47270	Cardiac resynchronization therapy implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E2401	Insufficient Information	A26	Insufficient Device Problem Information		
1758	2012		47270	Cardiac resynchronization therapy implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	F15	Recognized Device or Procedural Complication	A26	Insufficient Device Problem Information		
1759	2012		47270	Cardiac resynchronization therapy implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	F21	Unexpected Deterioration	A26	Insufficient Device Problem Information		
1760	2012		47270	Cardiac resynchronization therapy implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E2401	Insufficient Information	A26	Insufficient Device Problem Information		
1761	2012		47270	Cardiac resynchronization therapy implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E2401	Insufficient Information	A26	Insufficient Device Problem Information		
1762	2012		47270	Cardiac resynchronization therapy implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E2109	Implant Pain	A26	Insufficient Device Problem Information		
1763	2012		47270	Cardiac resynchronization therapy implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E2401	Insufficient Information	A1104	Computer Operating System Problem		
1764	2012		47270	Cardiac resynchronization therapy implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	F21	Unexpected Deterioration	A26	Insufficient Device Problem Information		
1765	2012		47270	Cardiac resynchronization therapy implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	F21	Unexpected Deterioration	A26	Insufficient Device Problem Information		
1766	2012		47270	Cardiac resynchronization therapy implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	F21	Unexpected Deterioration	A26	Insufficient Device Problem Information		
1767	2012				CT257	Implantable cardiac pacemakers	F15	Recognized Device or Procedural Complication	A26	Insufficient Device Problem Information		
1768	2012		47270	Cardiac resynchronization therapy implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	F15	Recognized Device or Procedural Complication	A26	Insufficient Device Problem Information		
1769	2012		47270	Cardiac resynchronization therapy implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	F21	Unexpected Deterioration	A26	Insufficient Device Problem Information		
1770	2012		47270	Cardiac resynchronization therapy implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E0506	Hemorrhage/Blood Loss/Bleeding	A26	Insufficient Device Problem Information		
1771	2011	All other reportable incidents			CT257	Implantable cardiac pacemakers	F1905	Device Revision or Replacement	A07	Electrical /Electronic Property Problem	24/11/2011	
									A0708	Power Problem	24/11/2011	
1772	2011	All other reportable incidents			CT257	Implantable cardiac pacemakers	F1905	Device Revision or Replacement	A07	Electrical /Electronic Property Problem	07/10/2011	
									A0708	Power Problem	07/10/2011	
1773	2011				CT257	Implantable cardiac pacemakers			A27	Appropriate Device Problem Term/Code not Available		
1774	2011		47270	Cardiac resynchronization therapy implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A26	Insufficient Device Problem Information		
1775	2012		47270	Cardiac resynchronization therapy implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A26	Insufficient Device Problem Information		
1776	2012	All other reportable incidents			CT257	Implantable cardiac pacemakers	F1905	Device Revision or Replacement	A07	Electrical /Electronic Property Problem		
									A0708	Power Problem		
1777	2011		47270	Cardiac resynchronization therapy implantable defibrillator	CT2866	Implantable cardiac electrical stimulators			A27	Appropriate Device Problem Term/Code not Available		
1778	2011		47270	Cardiac resynchronization therapy implantable defibrillator	CT2866	Implantable cardiac electrical stimulators			A26	Insufficient Device Problem Information		
1779	2011				CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A27	Appropriate Device Problem Term/Code not Available		
1780	2011	All other reportable incidents			CT257	Implantable cardiac pacemakers	F1905	Device Revision or Replacement	A07	Electrical /Electronic Property Problem		
									A0708	Power Problem		
1781	2012				CT257	Implantable cardiac pacemakers	F15	Recognized Device or Procedural Complication	A26	Insufficient Device Problem Information		
1782	2011		47270	Cardiac resynchronization therapy implantable defibrillator	CT2866	Implantable cardiac electrical stimulators			A26	Insufficient Device Problem Information		
1783	2012				CT257	Implantable cardiac pacemakers	F1905	Device Revision or Replacement	A07	Electrical /Electronic Property Problem		
									A0708	Power Problem		
1784	2012		47270	Cardiac resynchronization therapy implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E2401	Insufficient Information	A07	Electrical /Electronic Property Problem		
									A0709	Device Sensing Problem		

1785	2011	All other reportable incidents			CT257	Implantable cardiac pacemakers	F1905	Device Revision or Replacement	A09	Output Problem		
									A0907	No Device Output		
1786	2011	All other reportable incidents			CT257	Implantable cardiac pacemakers	F1905	Device Revision or Replacement	A07	Electrical /Electronic Property Problem	25/07/2011	
									A0708	Power Problem	25/07/2011	
1787	2012		47270	Cardiac resynchronization therapy implantable defibrillator	CT2866	Implantable cardiac electrical stimulators			A26	Insufficient Device Problem Information		
1788	2012	All other reportable incidents			CT257	Implantable cardiac pacemakers	F1905	Device Revision or Replacement	A07	Electrical /Electronic Property Problem	18/04/2012	
									A0708	Power Problem	18/04/2012	
1789	2011				CT257	Implantable cardiac pacemakers	F1905	Device Revision or Replacement	A07	Electrical /Electronic Property Problem		
									A0708	Power Problem		
1790	2011				CT257	Implantable cardiac pacemakers			A26	Insufficient Device Problem Information		
1791	2012				CT257	Implantable cardiac pacemakers	F1905	Device Revision or Replacement	A09	Output Problem		
									A090404	Output Above Specifications		
1792	2011				CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A07	Electrical /Electronic Property Problem		
									A0708	Power Problem		
1793	2012		47270	Cardiac resynchronization therapy implantable defibrillator	CT2866	Implantable cardiac electrical stimulators			A07	Electrical /Electronic Property Problem		
									A0708	Power Problem		
1794	2012				CT257	Implantable cardiac pacemakers	E2401	Insufficient Information	A26	Insufficient Device Problem Information		
1795	2012		47270	Cardiac resynchronization therapy implantable defibrillator	CT2866	Implantable cardiac electrical stimulators			A26	Insufficient Device Problem Information		
1796	2012		47270	Cardiac resynchronization therapy implantable defibrillator	CT2866	Implantable cardiac electrical stimulators			A26	Insufficient Device Problem Information		
1797	2012	All other reportable incidents			CT257	Implantable cardiac pacemakers	F1905	Device Revision or Replacement	A09	Output Problem	18/04/2012	
									A090404	Output Above Specifications	18/04/2012	
1798	2011		35223	Endocardial/interventricular septal pacing lead	CT1684	Cardiac pacemakers and associated devices			A27	Appropriate Device Problem Term/Code not Available		
1799	2012		47270	Cardiac resynchronization therapy implantable defibrillator	CT2866	Implantable cardiac electrical stimulators			A26	Insufficient Device Problem Information		
1800	2012				CT257	Implantable cardiac pacemakers	F1905	Device Revision or Replacement	A07	Electrical /Electronic Property Problem		
									A0708	Power Problem		
1801	2012		47270	Cardiac resynchronization therapy implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E0514	Thrombosis/Thrombus	A26	Insufficient Device Problem Information		
1802	2012		47270	Cardiac resynchronization therapy implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	F1905	Device Revision or Replacement	A07	Electrical /Electronic Property Problem		
									A0708	Power Problem		
1803	2012	All other reportable incidents			CT257	Implantable cardiac pacemakers	F1905	Device Revision or Replacement	A07	Electrical /Electronic Property Problem	31/08/2012	
									A0708	Power Problem	31/08/2012	
1804	2011				CT257	Implantable cardiac pacemakers			A11	Computer Software Problem		
									A1102	Application Program Problem		
1805	2012	All other reportable incidents			CT257	Implantable cardiac pacemakers	F1905	Device Revision or Replacement	A07	Electrical /Electronic Property Problem	23/04/2012	
									A0708	Power Problem	23/04/2012	
1806	2012		47270	Cardiac resynchronization therapy implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	F15	Recognized Device or Procedural Complication	A26	Insufficient Device Problem Information		
1807	2011				CT257	Implantable cardiac pacemakers	F1905	Device Revision or Replacement	A26	Insufficient Device Problem Information		
1808	2011				CT257	Implantable cardiac pacemakers	E2401	Insufficient Information	A07	Electrical /Electronic Property Problem		
									A0708	Power Problem		
1809	2011	All other reportable incidents			CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A09	Output Problem		
									A0907	No Device Output		
1810	2011	All other reportable incidents			CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A1304	Telemetry Discrepancy		
									A26	Insufficient Device Problem Information		
1811	2011	All other reportable incidents			CT257	Implantable cardiac pacemakers	F1905	Device Revision or Replacement	A1304	Telemetry Discrepancy	13/07/2011	
									A26	Insufficient Device Problem Information	13/07/2011	
1812	2012		47270	Cardiac resynchronization therapy implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E1906	Unspecified Infection	A26	Insufficient Device Problem Information		
							E2104	Electric Shock	A26	Insufficient Device Problem Information		
							E2401	Insufficient Information	A26	Insufficient Device Problem Information		
							F1905	Device Revision or Replacement	A26	Insufficient Device Problem Information		
1813	2012	All other reportable incidents	35039	Epicardial pacing lead	CT1684	Cardiac pacemakers and associated devices	F1905	Device Revision or Replacement	A07	Electrical /Electronic Property Problem	16/08/2012	
									A0709	Device Sensing Problem	16/08/2012	
1814	2011		47270	Cardiac resynchronization therapy implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A07	Electrical /Electronic Property Problem		
									A0709	Device Sensing Problem		
1815	2012				CT257	Implantable cardiac pacemakers			A07	Electrical /Electronic Property Problem		
									A0708	Power Problem		
1816	2012		47270	Cardiac resynchronization therapy implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E2401	Insufficient Information	A11	Computer Software Problem		
									A1102	Application Program Problem		
									A1104	Computer Operating System Problem		
1817	2011		35223	Endocardial/interventricular septal pacing lead	CT1684	Cardiac pacemakers and associated devices			A26	Insufficient Device Problem Information		
1818	2012				CT257	Implantable cardiac pacemakers	F1905	Device Revision or Replacement	A07	Electrical /Electronic Property Problem		
									A0708	Power Problem		
1819	2012				CT257	Implantable cardiac pacemakers	F15	Recognized Device or Procedural Complication	A26	Insufficient Device Problem Information		
1820	2012		47270	Cardiac resynchronization therapy implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	F15	Recognized Device or Procedural Complication	A26	Insufficient Device Problem Information		
1821	2012		47270	Cardiac resynchronization therapy implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	F15	Recognized Device or Procedural Complication	A26	Insufficient Device Problem Information		
1822	2011		35223	Endocardial/interventricular septal pacing lead	CT1684	Cardiac pacemakers and associated devices			A27	Appropriate Device Problem Term/Code not Available		
1823	2011		35223	Endocardial/interventricular septal pacing lead	CT1684	Cardiac pacemakers and associated devices			A04	Material Integrity Problem		
									A0405	Degraded		
1824	2011				CT257	Implantable cardiac pacemakers			A1304	Telemetry Discrepancy		
									A26	Insufficient Device Problem Information		
1825	2012		47270	Cardiac resynchronization therapy implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E1906	Unspecified Infection	A26	Insufficient Device Problem Information		
1826	2012		47270	Cardiac resynchronization therapy implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	F1905	Device Revision or Replacement	A26	Insufficient Device Problem Information		
1827	2012		47270	Cardiac resynchronization therapy implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	F21	Unexpected Deterioration	A26	Insufficient Device Problem Information		
1828	2012		47270	Cardiac resynchronization therapy implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E2401	Insufficient Information	A26	Insufficient Device Problem Information		
1829	2012		47270	Cardiac resynchronization therapy implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E2007	Fall	A26	Insufficient Device Problem Information		
1830	2012		47270	Cardiac resynchronization therapy implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E0402	Hypersensitivity/Allergic Reaction	A26	Insufficient Device Problem Information		

1831	2012		47270	Cardiac resynchronization therapy implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E2401	Insufficient Information	A26	Insufficient Device Problem Information		
1832	2012		47270	Cardiac resynchronization therapy implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E2401	Insufficient Information	A26	Insufficient Device Problem Information		
1833	2012		47270	Cardiac resynchronization therapy implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	F15	Recognized Device or Procedural Complication	A26	Insufficient Device Problem Information		
1834	2012		47270	Cardiac resynchronization therapy implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E2401	Insufficient Information	A07	Electrical /Electronic Property Problem		
									A0708	Power Problem		
									A09	Output Problem		
									A090404	Output Above Specifications		
1835	2011		47270	Cardiac resynchronization therapy implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E2401	Insufficient Information	A26	Insufficient Device Problem Information		
1836	2012		47270	Cardiac resynchronization therapy implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	F15	Recognized Device or Procedural Complication	A26	Insufficient Device Problem Information		
1837	2011				CT257	Implantable cardiac pacemakers	E2401	Insufficient Information	A07	Electrical /Electronic Property Problem		
									A0708	Power Problem		
1838	2012		47270	Cardiac resynchronization therapy implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	F21	Unexpected Deterioration	A26	Insufficient Device Problem Information		
1839	2012		47270	Cardiac resynchronization therapy implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E0402	Hypersensitivity/Allergic Reaction	A26	Insufficient Device Problem Information		
1840	2012		47270	Cardiac resynchronization therapy implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	F15	Recognized Device or Procedural Complication	A26	Insufficient Device Problem Information		
1841	2012		47270	Cardiac resynchronization therapy implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	F14	Prolonged Episode of Care	A26	Insufficient Device Problem Information		
1842	2012	All other reportable incidents			CT257	Implantable cardiac pacemakers	F1905	Device Revision or Replacement	A07	Electrical /Electronic Property Problem	08/11/2011	
									A0708	Power Problem	08/11/2011	
1843	2012		47270	Cardiac resynchronization therapy implantable defibrillator	CT2866	Implantable cardiac electrical stimulators			A26	Insufficient Device Problem Information		
1844	2012		47270	Cardiac resynchronization therapy implantable defibrillator	CT2866	Implantable cardiac electrical stimulators			A26	Insufficient Device Problem Information		
1845	2012		47270	Cardiac resynchronization therapy implantable defibrillator	CT2866	Implantable cardiac electrical stimulators			A26	Insufficient Device Problem Information		
1846	2012	All other reportable incidents			CT257	Implantable cardiac pacemakers	F1905	Device Revision or Replacement	A19	Environmental Compatibility Problem	05/10/2011	
									A191002	Loss of Power	05/10/2011	
1847	2011		47270	Cardiac resynchronization therapy implantable defibrillator	CT2866	Implantable cardiac electrical stimulators			A27	Appropriate Device Problem Term/Code not Available		
1848	2011		47270	Cardiac resynchronization therapy implantable defibrillator	CT2866	Implantable cardiac electrical stimulators			A27	Appropriate Device Problem Term/Code not Available		
1849	2011		47270	Cardiac resynchronization therapy implantable defibrillator	CT2866	Implantable cardiac electrical stimulators			A27	Appropriate Device Problem Term/Code not Available		
1850	2012	All other reportable incidents			CT257	Implantable cardiac pacemakers	F1905	Device Revision or Replacement	A09	Output Problem	27/08/2012	
									A0907	No Device Output	27/08/2012	
1851	2012		47270	Cardiac resynchronization therapy implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A26	Insufficient Device Problem Information		
1852	2012		47270	Cardiac resynchronization therapy implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	F21	Unexpected Deterioration	A26	Insufficient Device Problem Information		
1853	2012		47270	Cardiac resynchronization therapy implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E0402	Hypersensitivity/Allergic Reaction	A26	Insufficient Device Problem Information		
1854	2011		47270	Cardiac resynchronization therapy implantable defibrillator	CT2866	Implantable cardiac electrical stimulators			A26	Insufficient Device Problem Information		
1855	2012		47270	Cardiac resynchronization therapy implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	F21	Unexpected Deterioration	A26	Insufficient Device Problem Information		
1856	2012		47270	Cardiac resynchronization therapy implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	F21	Unexpected Deterioration	A26	Insufficient Device Problem Information		
1857	2011				CT257	Implantable cardiac pacemakers			A27	Appropriate Device Problem Term/Code not Available		
1858	2012		47270	Cardiac resynchronization therapy implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	F15	Recognized Device or Procedural Complication	A26	Insufficient Device Problem Information		
1859	2012		47270	Cardiac resynchronization therapy implantable defibrillator	CT2866	Implantable cardiac electrical stimulators			A26	Insufficient Device Problem Information		
1860	2012		47270	Cardiac resynchronization therapy implantable defibrillator	CT2866	Implantable cardiac electrical stimulators			A26	Insufficient Device Problem Information		
1861	2012		47270	Cardiac resynchronization therapy implantable defibrillator	CT2866	Implantable cardiac electrical stimulators			A26	Insufficient Device Problem Information		
1862	2012		47270	Cardiac resynchronization therapy implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	F15	Recognized Device or Procedural Complication	A26	Insufficient Device Problem Information		
1863	2012				CT257	Implantable cardiac pacemakers	E2401	Insufficient Information	A07	Electrical /Electronic Property Problem		
									A0708	Power Problem		
1864	2012				CT257	Implantable cardiac pacemakers	F1905	Device Revision or Replacement	A07	Electrical /Electronic Property Problem		
									A0708	Power Problem		
1865	2012				CT257	Implantable cardiac pacemakers	F15	Recognized Device or Procedural Complication	A26	Insufficient Device Problem Information		
1866	2012				CT257	Implantable cardiac pacemakers	E2401	Insufficient Information	A26	Insufficient Device Problem Information		
1867	2012		47270	Cardiac resynchronization therapy implantable defibrillator	CT2866	Implantable cardiac electrical stimulators			A26	Insufficient Device Problem Information		
1868	2012		47270	Cardiac resynchronization therapy implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	F1905	Device Revision or Replacement	A07	Electrical /Electronic Property Problem		
									A0709	Device Sensing Problem		
1869	2011				CT257	Implantable cardiac pacemakers	F1905	Device Revision or Replacement	A26	Insufficient Device Problem Information		
1870	2011				CT257	Implantable cardiac pacemakers	F1905	Device Revision or Replacement	A07	Electrical /Electronic Property Problem		
									A0709	Device Sensing Problem		
1871	2012	All other reportable incidents			CT257	Implantable cardiac pacemakers	F1905	Device Revision or Replacement	A07	Electrical /Electronic Property Problem		
									A0708	Power Problem		
1872	2011				CT257	Implantable cardiac pacemakers			A27	Appropriate Device Problem Term/Code not Available		
1873	2011				CT257	Implantable cardiac pacemakers			A27	Appropriate Device Problem Term/Code not Available		
1874	2011				CT257	Implantable cardiac pacemakers			A27	Appropriate Device Problem Term/Code not Available		
1875	2011				CT257	Implantable cardiac pacemakers			A27	Appropriate Device Problem Term/Code not Available		
1876	2012		47270	Cardiac resynchronization therapy implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	F1905	Device Revision or Replacement	A23	Use of Device Problem		
1877	2012		47270	Cardiac resynchronization therapy implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	F21	Unexpected Deterioration	A05	Mechanical Problem		
									A0501	Detachment of Device or Device Component		
									A09	Output Problem		
									A0908	Incorrect, Inadequate or Imprecise Result or Readings		

1878	2012		47270	Cardiac resynchronization therapy implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A07	Electrical /Electronic Property Problem		
									A0709	Device Sensing Problem		
1879	2011		47270	Cardiac resynchronization therapy implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E2401	Insufficient Information	A26	Insufficient Device Problem Information		
1880	2011		47270	Cardiac resynchronization therapy implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E2401	Insufficient Information	A26	Insufficient Device Problem Information		
1881	2012				CT257	Implantable cardiac pacemakers	F1905	Device Revision or Replacement	A07	Electrical /Electronic Property Problem		
									A0709	Device Sensing Problem		
									A09	Output Problem		
									A0907	No Device Output		
1882	2012	All other reportable incidents			CT1684	Cardiac pacemakers and associated devices			A07	Electrical /Electronic Property Problem		
									A0709	Device Sensing Problem		
									A09	Output Problem		
									A0908	Incorrect, Inadequate or Imprecise Result or Readings		
1883	2012		47270	Cardiac resynchronization therapy implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E2401	Insufficient Information	A10	Temperature Problem		
									A1006	Thermal Decomposition of Device		
1884	2011	All other reportable incidents	47270	Cardiac resynchronization therapy implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	F1905	Device Revision or Replacement	A07	Electrical /Electronic Property Problem	08/08/2011	
									A0708	Power Problem	08/08/2011	
1885	2011		47263	Cardiac resynchronization therapy implantable pacemaker	CT1684	Cardiac pacemakers and associated devices	F1905	Device Revision or Replacement	A07	Electrical /Electronic Property Problem		
									A0709	Device Sensing Problem		
					CT1685	Cardiac pacemakers	F1905	Device Revision or Replacement	A07	Electrical /Electronic Property Problem		
									A0709	Device Sensing Problem		
					CT257	Implantable cardiac pacemakers	F1905	Device Revision or Replacement	A07	Electrical /Electronic Property Problem		
									A0709	Device Sensing Problem		
					CT2866	Implantable cardiac electrical stimulators	F1905	Device Revision or Replacement	A07	Electrical /Electronic Property Problem		
									A0709	Device Sensing Problem		
1886	2012		47270	Cardiac resynchronization therapy implantable defibrillator	CT2866	Implantable cardiac electrical stimulators			A26	Insufficient Device Problem Information		
1887	2011		47270	Cardiac resynchronization therapy implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E2401	Insufficient Information	A26	Insufficient Device Problem Information		
1888	2011	All other reportable incidents			CT257	Implantable cardiac pacemakers	F1905	Device Revision or Replacement	A1304	Telemetry Discrepancy		
									A26	Insufficient Device Problem Information		
1889	2011		47270	Cardiac resynchronization therapy implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	F1905	Device Revision or Replacement	A26	Insufficient Device Problem Information		
1890	2012				CT257	Implantable cardiac pacemakers	F15	Recognized Device or Procedural Complication	A26	Insufficient Device Problem Information		
1891	2012				CT257	Implantable cardiac pacemakers	F1905	Device Revision or Replacement	A07	Electrical /Electronic Property Problem		
									A0708	Power Problem		
									A09	Output Problem		
									A0907	No Device Output		
									A13	Communication or Transmission Problem		
									A26	Insufficient Device Problem Information		
1892	2012		47270	Cardiac resynchronization therapy implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	F15	Recognized Device or Procedural Complication	A26	Insufficient Device Problem Information		
1893	2012		47270	Cardiac resynchronization therapy implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	F15	Recognized Device or Procedural Complication	A26	Insufficient Device Problem Information		
1894	2012		47270	Cardiac resynchronization therapy implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	F15	Recognized Device or Procedural Complication	A26	Insufficient Device Problem Information		
1895	2012		47270	Cardiac resynchronization therapy implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A26	Insufficient Device Problem Information		
1896	2012		47270	Cardiac resynchronization therapy implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	F15	Recognized Device or Procedural Complication	A26	Insufficient Device Problem Information		
1897	2012		47270	Cardiac resynchronization therapy implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	F15	Recognized Device or Procedural Complication	A26	Insufficient Device Problem Information		
1898	2012		47263	Cardiac resynchronization therapy implantable pacemaker	CT1684	Cardiac pacemakers and associated devices	F1905	Device Revision or Replacement	A07	Electrical /Electronic Property Problem		
									A0708	Power Problem		
					CT1685	Cardiac pacemakers	F1905	Device Revision or Replacement	A07	Electrical /Electronic Property Problem		
									A0708	Power Problem		
					CT257	Implantable cardiac pacemakers	F1905	Device Revision or Replacement	A07	Electrical /Electronic Property Problem		
									A0708	Power Problem		
					CT2866	Implantable cardiac electrical stimulators	F1905	Device Revision or Replacement	A07	Electrical /Electronic Property Problem		
									A0708	Power Problem		
1899	2012		47270	Cardiac resynchronization therapy implantable defibrillator	CT2866	Implantable cardiac electrical stimulators			A26	Insufficient Device Problem Information		
1900	2012		47270	Cardiac resynchronization therapy implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	F21	Unexpected Deterioration	A26	Insufficient Device Problem Information		
1901	2012	All other reportable incidents			CT257	Implantable cardiac pacemakers	F1905	Device Revision or Replacement	A07	Electrical /Electronic Property Problem		
									A0708	Power Problem		
1902	2011		47270	Cardiac resynchronization therapy implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E2401	Insufficient Information	A26	Insufficient Device Problem Information		
1903	2012				CT257	Implantable cardiac pacemakers	E0402	Hypersensitivity/Allergic Reaction	A26	Insufficient Device Problem Information		
1904	2012		47270	Cardiac resynchronization therapy implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	F14	Prolonged Episode of Care	A26	Insufficient Device Problem Information		
1905	2012				CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A07	Electrical /Electronic Property Problem		
									A0708	Power Problem		
1906	2011		47270	Cardiac resynchronization therapy implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	F21	Unexpected Deterioration	A26	Insufficient Device Problem Information		
1907	2011		47270	Cardiac resynchronization therapy implantable defibrillator	CT2866	Implantable cardiac electrical stimulators			A27	Appropriate Device Problem Term/Code not Available		
1908	2012				CT257	Implantable cardiac pacemakers	F1905	Device Revision or Replacement	A07	Electrical /Electronic Property Problem		
									A0708	Power Problem		
1909	2012	All other reportable incidents	47270	Cardiac resynchronization therapy implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	F1905	Device Revision or Replacement	A07	Electrical /Electronic Property Problem	14/05/2012	
									A0708	Power Problem	14/05/2012	
1910	2011				CT257	Implantable cardiac pacemakers	F1905	Device Revision or Replacement	A07	Electrical /Electronic Property Problem		
									A0708	Power Problem		
1911	2011				CT257	Implantable cardiac pacemakers	F1905	Device Revision or Replacement	A07	Electrical /Electronic Property Problem		
									A0708	Power Problem		
1912	2012				CT257	Implantable cardiac pacemakers	F1905	Device Revision or Replacement	A0104	Migration or Expulsion of Device		
1913	2012		47270	Cardiac resynchronization therapy implantable defibrillator	CT2866	Implantable cardiac electrical stimulators			A26	Insufficient Device Problem Information		
1914	2012	All other reportable incidents			CT257	Implantable cardiac pacemakers	F1905	Device Revision or Replacement	A09	Output Problem	13/09/2012	
									A0907	No Device Output	13/09/2012	

									A13	Communication or Transmission Problem	13/09/2012	
									A26	Insufficient Device Problem Information	13/09/2012	
1915	2012		47270	Cardiac resynchronization therapy implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	F15	Recognized Device or Procedural Complication	A26	Insufficient Device Problem Information		
1916	2012				CT257	Implantable cardiac pacemakers	E2401	Insufficient Information	A26	Insufficient Device Problem Information		
1917	2012		47270	Cardiac resynchronization therapy implantable defibrillator	CT2866	Implantable cardiac electrical stimulators			A09	Output Problem		
									A090404	Output Above Specifications		
									A12	Connection Problem		
									A1204	Failure to Disconnect		
									A15	Activation, Positioning or Separation Problem		
									A150301	Separation Failure		
1918	2011		47270	Cardiac resynchronization therapy implantable defibrillator	CT2866	Implantable cardiac electrical stimulators			A26	Insufficient Device Problem Information		
1919	2011				CT257	Implantable cardiac pacemakers			A27	Appropriate Device Problem Term/Code not Available		
1920	2012				CT257	Implantable cardiac pacemakers	F15	Recognized Device or Procedural Complication	A26	Insufficient Device Problem Information		
1921	2012	All other reportable incidents			CT257	Implantable cardiac pacemakers	F1905	Device Revision or Replacement	A07	Electrical /Electronic Property Problem	16/05/2012	
									A0708	Power Problem	16/05/2012	
1922	2011	All other reportable incidents			CT257	Implantable cardiac pacemakers			A07	Electrical /Electronic Property Problem	14/06/2011	
									A0708	Power Problem	14/06/2011	
1923	2011		35223	Endocardial/interventricular septal pacing lead	CT1684	Cardiac pacemakers and associated devices	F21	Unexpected Deterioration	A07	Electrical /Electronic Property Problem		
									A0709	Device Sensing Problem		
1924	2011		47270	Cardiac resynchronization therapy implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E2401	Insufficient Information	A26	Insufficient Device Problem Information		
1925	2012		47270	Cardiac resynchronization therapy implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	F14	Prolonged Episode of Care	A09	Output Problem		
									A090404	Output Above Specifications		
1926	2012		47270	Cardiac resynchronization therapy implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E0402	Hypersensitivity/Allergic Reaction	A01	Patient Device Interaction Problem		
									A17	Compatibility Problem		
1927	2012		47270	Cardiac resynchronization therapy implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	F14	Prolonged Episode of Care	A26	Insufficient Device Problem Information		
1928	2012		47270	Cardiac resynchronization therapy implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	F21	Unexpected Deterioration	A26	Insufficient Device Problem Information		
1929	2012		47270	Cardiac resynchronization therapy implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	F21	Unexpected Deterioration	A26	Insufficient Device Problem Information		
1930	2012		47270	Cardiac resynchronization therapy implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	F21	Unexpected Deterioration	A26	Insufficient Device Problem Information		
1931	2012		47270	Cardiac resynchronization therapy implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	F21	Unexpected Deterioration	A26	Insufficient Device Problem Information		
1932	2012		47270	Cardiac resynchronization therapy implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	F21	Unexpected Deterioration	A26	Insufficient Device Problem Information		
1933	2012	All other reportable incidents			CT257	Implantable cardiac pacemakers	F1905	Device Revision or Replacement	A07	Electrical /Electronic Property Problem	16/05/2012	
									A0708	Power Problem	16/05/2012	
1934	2011				CT257	Implantable cardiac pacemakers	E2401	Insufficient Information	A11	Computer Software Problem		
									A1102	Application Program Problem		
									A15	Activation, Positioning or Separation Problem		
									A150102	Difficult or Delayed Activation		
1935	2012	All other reportable incidents	35223	Endocardial/interventricular septal pacing lead	CT1684	Cardiac pacemakers and associated devices	F1905	Device Revision or Replacement	A07	Electrical /Electronic Property Problem	17/07/2012	
									A0709	Device Sensing Problem	17/07/2012	
1936	2012	All other reportable incidents			CT257	Implantable cardiac pacemakers	F1905	Device Revision or Replacement	A07	Electrical /Electronic Property Problem		
									A0708	Power Problem		
1937	2011				CT257	Implantable cardiac pacemakers	F1905	Device Revision or Replacement	A11	Computer Software Problem		
									A1102	Application Program Problem		
1938	2011				CT257	Implantable cardiac pacemakers	E2401	Insufficient Information	A07	Electrical /Electronic Property Problem		
									A0708	Power Problem		
1939	2011				CT257	Implantable cardiac pacemakers	F1905	Device Revision or Replacement	A07	Electrical /Electronic Property Problem		
									A0708	Power Problem		
1940	2012				CT257	Implantable cardiac pacemakers	F15	Recognized Device or Procedural Complication	A26	Insufficient Device Problem Information		
1941	2011				CT257	Implantable cardiac pacemakers	F1905	Device Revision or Replacement	A07	Electrical /Electronic Property Problem		
									A0709	Device Sensing Problem		
1942	2012		35223	Endocardial/interventricular septal pacing lead	CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A26	Insufficient Device Problem Information		
1943	2012		35223	Endocardial/interventricular septal pacing lead	CT1684	Cardiac pacemakers and associated devices			A07	Electrical /Electronic Property Problem		
									A0709	Device Sensing Problem		
1944	2012				CT257	Implantable cardiac pacemakers	F1905	Device Revision or Replacement	A07	Electrical /Electronic Property Problem		
									A0708	Power Problem		
									A0709	Device Sensing Problem		
1945	2012				CT257	Implantable cardiac pacemakers	F15	Recognized Device or Procedural Complication	A26	Insufficient Device Problem Information		
1946	2011				CT257	Implantable cardiac pacemakers			A09	Output Problem		
									A090404	Output Above Specifications		
1947	2012		47270	Cardiac resynchronization therapy implantable defibrillator	CT2866	Implantable cardiac electrical stimulators			A26	Insufficient Device Problem Information		
1948	2012		47270	Cardiac resynchronization therapy implantable defibrillator	CT2866	Implantable cardiac electrical stimulators			A26	Insufficient Device Problem Information		
1949	2012				CT257	Implantable cardiac pacemakers			A26	Insufficient Device Problem Information		
1950	2012	All other reportable incidents			CT257	Implantable cardiac pacemakers	F1905	Device Revision or Replacement	A07	Electrical /Electronic Property Problem	16/10/2012	
									A0708	Power Problem	16/10/2012	
1951	2012	All other reportable incidents			CT257	Implantable cardiac pacemakers	F1905	Device Revision or Replacement	A07	Electrical /Electronic Property Problem	10/10/2012	
									A0708	Power Problem	10/10/2012	
1952	2012	All other reportable incidents			CT257	Implantable cardiac pacemakers	F1905	Device Revision or Replacement	A07	Electrical /Electronic Property Problem	20/10/2012	
									A0708	Power Problem	20/10/2012	
1953	2012	All other reportable incidents	47270	Cardiac resynchronization therapy implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	F1905	Device Revision or Replacement	A07	Electrical /Electronic Property Problem	20/09/2012	
									A0708	Power Problem	20/09/2012	
1954	2012		47270	Cardiac resynchronization therapy implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	F15	Recognized Device or Procedural Complication	A26	Insufficient Device Problem Information		
1955	2011	All other reportable incidents			CT257	Implantable cardiac pacemakers	F1905	Device Revision or Replacement	A07	Electrical /Electronic Property Problem	08/03/2011	
									A0708	Power Problem	08/03/2011	
1956	2012				CT257	Implantable cardiac pacemakers	F1905	Device Revision or Replacement	A07	Electrical /Electronic Property Problem		
									A0708	Power Problem		
1957	2012		47270	Cardiac resynchronization therapy implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E2104	Electric Shock	A09	Output Problem		
									A090404	Output Above Specifications		
1958	2011				CT257	Implantable cardiac pacemakers	F1905	Device Revision or Replacement	A1304	Telemetry Discrepancy		

									A26	Insufficient Device Problem Information		
1959	2011	All other reportable incidents			CT257	Implantable cardiac pacemakers	F1905	Device Revision or Replacement	A04	Material Integrity Problem		
									A0405	Degraded		
1960	2011		47263	Cardiac resynchronization therapy implantable pacemaker	CT1684	Cardiac pacemakers and associated devices	F1905	Device Revision or Replacement	A07	Electrical /Electronic Property Problem		
									A0708	Power Problem		
					CT1685	Cardiac pacemakers	F1905	Device Revision or Replacement	A07	Electrical /Electronic Property Problem		
									A0708	Power Problem		
					CT257	Implantable cardiac pacemakers	F1905	Device Revision or Replacement	A07	Electrical /Electronic Property Problem		
									A0708	Power Problem		
					CT2866	Implantable cardiac electrical stimulators	F1905	Device Revision or Replacement	A07	Electrical /Electronic Property Problem		
									A0708	Power Problem		
1961	2011				CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A07	Electrical /Electronic Property Problem		
									A0708	Power Problem		
1962	2012				CT257	Implantable cardiac pacemakers	F1905	Device Revision or Replacement	A09	Output Problem		
									A0907	No Device Output		
1963	2012		47270	Cardiac resynchronization therapy implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A07	Electrical /Electronic Property Problem		
									A0709	Device Sensing Problem		
1964	2012				CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A07	Electrical /Electronic Property Problem		
									A0708	Power Problem		
1965	2012		35223	Endocardial/interventricular septal pacing lead	CT1684	Cardiac pacemakers and associated devices	F1905	Device Revision or Replacement	A04	Material Integrity Problem		
									A0405	Degraded		
1966	2012		47270	Cardiac resynchronization therapy implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	F15	Recognized Device or Procedural Complication	A26	Insufficient Device Problem Information		
1967	2012		47270	Cardiac resynchronization therapy implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A26	Insufficient Device Problem Information		
1968	2012		47270	Cardiac resynchronization therapy implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	F21	Unexpected Deterioration	A26	Insufficient Device Problem Information		
1969	2012		47270	Cardiac resynchronization therapy implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A26	Insufficient Device Problem Information		
1970	2012		47270	Cardiac resynchronization therapy implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E0402	Hypersensitivity/Allergic Reaction	A26	Insufficient Device Problem Information		
1971	2012				CT257	Implantable cardiac pacemakers	F1905	Device Revision or Replacement	A07	Electrical /Electronic Property Problem		
									A0708	Power Problem		
									A0709	Device Sensing Problem		
1972	2012				CT257	Implantable cardiac pacemakers			A26	Insufficient Device Problem Information		
1973	2012		35223	Endocardial/interventricular septal pacing lead	CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A07	Electrical /Electronic Property Problem		
									A0709	Device Sensing Problem		
1974	2012		47270	Cardiac resynchronization therapy implantable defibrillator	CT2866	Implantable cardiac electrical stimulators			A26	Insufficient Device Problem Information		
1975	2011				CT257	Implantable cardiac pacemakers	F1905	Device Revision or Replacement	A07	Electrical /Electronic Property Problem		
									A0708	Power Problem		
1976	2012				CT257	Implantable cardiac pacemakers	F1905	Device Revision or Replacement	A07	Electrical /Electronic Property Problem		
									A0708	Power Problem		
1977	2012	All other reportable incidents			CT257	Implantable cardiac pacemakers	E2401	Insufficient Information	A07	Electrical /Electronic Property Problem		
									A0708	Power Problem		
1978	2011		47270	Cardiac resynchronization therapy implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A07	Electrical /Electronic Property Problem		
									A0709	Device Sensing Problem		
1979	2012				CT257	Implantable cardiac pacemakers	E2401	Insufficient Information	A26	Insufficient Device Problem Information		
1980	2011		47270	Cardiac resynchronization therapy implantable defibrillator	CT2866	Implantable cardiac electrical stimulators			A26	Insufficient Device Problem Information		
1981	2012				CT257	Implantable cardiac pacemakers	F1905	Device Revision or Replacement	A13	Communication or Transmission Problem		
									A26	Insufficient Device Problem Information		
1982	2011	All other reportable incidents	47270	Cardiac resynchronization therapy implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	F1905	Device Revision or Replacement	A07	Electrical /Electronic Property Problem	16/03/2011	
									A0708	Power Problem	16/03/2011	
1983	2012				CT257	Implantable cardiac pacemakers			A07	Electrical /Electronic Property Problem		
									A0709	Device Sensing Problem		
									A1304	Telemetry Discrepancy		
									A26	Insufficient Device Problem Information		
1984	2011				CT257	Implantable cardiac pacemakers			A09	Output Problem		
									A0907	No Device Output		
1985	2011		47263	Cardiac resynchronization therapy implantable pacemaker	CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A11	Computer Software Problem		
									A1102	Application Program Problem		
					CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A11	Computer Software Problem		
									A1102	Application Program Problem		
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A11	Computer Software Problem		
									A1102	Application Program Problem		
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A11	Computer Software Problem		
									A1102	Application Program Problem		
1986	2011		35223	Endocardial/interventricular septal pacing lead	CT1684	Cardiac pacemakers and associated devices	E2401	Insufficient Information	A09	Output Problem		
									A090404	Output Above Specifications		
1987	2012		47270	Cardiac resynchronization therapy implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	F15	Recognized Device or Procedural Complication	A26	Insufficient Device Problem Information		
1988	2012		47270	Cardiac resynchronization therapy implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	F15	Recognized Device or Procedural Complication	A26	Insufficient Device Problem Information		
1989	2012		35223	Endocardial/interventricular septal pacing lead	CT1684	Cardiac pacemakers and associated devices	F1905	Device Revision or Replacement	A07	Electrical /Electronic Property Problem		
									A0709	Device Sensing Problem		
1990	2012				CT257	Implantable cardiac pacemakers	F15	Recognized Device or Procedural Complication	A26	Insufficient Device Problem Information		
1991	2012				CT257	Implantable cardiac pacemakers	F21	Unexpected Deterioration	A26	Insufficient Device Problem Information		
1992	2012				CT257	Implantable cardiac pacemakers	E0514	Thrombosis/Thrombus	A07	Electrical /Electronic Property Problem		
									A0709	Device Sensing Problem		
									A26	Insufficient Device Problem Information		
							E1906	Unspecified Infection	A07	Electrical /Electronic Property Problem		
									A0709	Device Sensing Problem		
									A26	Insufficient Device Problem Information		
1993	2011	All other reportable incidents	35223	Endocardial/interventricular septal pacing lead	CT1684	Cardiac pacemakers and associated devices	F1905	Device Revision or Replacement	A07	Electrical /Electronic Property Problem		
									A0709	Device Sensing Problem		
1994	2011		47270	Cardiac resynchronization therapy implantable defibrillator	CT2866	Implantable cardiac electrical stimulators			A26	Insufficient Device Problem Information		
1995	2012				CT257	Implantable cardiac pacemakers	F1905	Device Revision or Replacement	A07	Electrical /Electronic Property Problem		
									A0708	Power Problem		

1996	2012		47270	Cardiac resynchronization therapy implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E1906	Unspecified Infection	A26	Insufficient Device Problem Information		
1997	2012		47270	Cardiac resynchronization therapy implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	F21	Unexpected Deterioration	A26	Insufficient Device Problem Information		
1998	2012		47270	Cardiac resynchronization therapy implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	F21	Unexpected Deterioration	A26	Insufficient Device Problem Information		
1999	2012		47270	Cardiac resynchronization therapy implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	F21	Unexpected Deterioration	A26	Insufficient Device Problem Information		
2000	2012		47270	Cardiac resynchronization therapy implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E2401	Insufficient Information	A26	Insufficient Device Problem Information		
2001	2011				CT257	Implantable cardiac pacemakers			A07	Electrical /Electronic Property Problem		
									A0708	Power Problem		
2002	2012		47270	Cardiac resynchronization therapy implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	F14	Prolonged Episode of Care	A26	Insufficient Device Problem Information		
2003	2011		47270	Cardiac resynchronization therapy implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E0402	Hypersensitivity/Allergic Reaction	A01	Patient Device Interaction Problem		
									A17	Compatibility Problem		
							F15	Recognized Device or Procedural Complication	A01	Patient Device Interaction Problem		
									A17	Compatibility Problem		
2004	2011		47270	Cardiac resynchronization therapy implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	F15	Recognized Device or Procedural Complication	A23	Use of Device Problem		
2005	2012		36104	Implantable pacemaker analyser	CT1684	Cardiac pacemakers and associated devices			A09	Output Problem		
									A0907	No Device Output		
2006	2011	All other reportable incidents			CT257	Implantable cardiac pacemakers	F1905	Device Revision or Replacement	A07	Electrical /Electronic Property Problem	18/04/2011	
									A0708	Power Problem	18/04/2011	
2007	2012				CT257	Implantable cardiac pacemakers	E2401	Insufficient Information	A15	Activation, Positioning or Separation Problem		
									A150301	Separation Failure		
2008	2012				CT257	Implantable cardiac pacemakers	F1905	Device Revision or Replacement	A07	Electrical /Electronic Property Problem		
									A0708	Power Problem		
2009	2012				CT257	Implantable cardiac pacemakers	F1905	Device Revision or Replacement	A07	Electrical /Electronic Property Problem		
									A0708	Power Problem		
2010	2012				CT257	Implantable cardiac pacemakers	E2401	Insufficient Information	A26	Insufficient Device Problem Information		
2011	2012				CT257	Implantable cardiac pacemakers	E2401	Insufficient Information	A26	Insufficient Device Problem Information		
2012	2012				CT257	Implantable cardiac pacemakers	E2401	Insufficient Information	A07	Electrical /Electronic Property Problem		
									A0709	Device Sensing Problem		
2013	2012				CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A13	Communication or Transmission Problem		
									A26	Insufficient Device Problem Information		
2014	2011		47270	Cardiac resynchronization therapy implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	F21	Unexpected Deterioration	A26	Insufficient Device Problem Information		
2015	2011		47270	Cardiac resynchronization therapy implantable defibrillator	CT2866	Implantable cardiac electrical stimulators			A09	Output Problem		
									A090404	Output Above Specifications		
2016	2011		47270	Cardiac resynchronization therapy implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	F14	Prolonged Episode of Care	A26	Insufficient Device Problem Information		
2017	2011		47270	Cardiac resynchronization therapy implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A07	Electrical /Electronic Property Problem		
									A0709	Device Sensing Problem		
2018	2012				CT257	Implantable cardiac pacemakers	F1905	Device Revision or Replacement	A07	Electrical /Electronic Property Problem		
									A0708	Power Problem		
2019	2012				CT257	Implantable cardiac pacemakers	E2401	Insufficient Information	A26	Insufficient Device Problem Information		
2020	2011	All other reportable incidents			CT257	Implantable cardiac pacemakers			A07	Electrical /Electronic Property Problem		
									A0708	Power Problem		
2021	2011				CT257	Implantable cardiac pacemakers	E2401	Insufficient Information	A11	Computer Software Problem		
									A1102	Application Program Problem		
2022	2012	All other reportable incidents			CT257	Implantable cardiac pacemakers	F1905	Device Revision or Replacement	A07	Electrical /Electronic Property Problem		
									A0708	Power Problem		
2023	2012	All other reportable incidents			CT257	Implantable cardiac pacemakers	F1905	Device Revision or Replacement	A09	Output Problem	31/10/2012	
									A0907	No Device Output	31/10/2012	
2024	2012	Death			CT1684	Cardiac pacemakers and associated devices	F21	Unexpected Deterioration	A09	Output Problem		
									A0908	Incorrect, Inadequate or Imprecise Result or Readings		
									A11	Computer Software Problem		
									A1102	Application Program Problem		
2025	2012	All other reportable incidents			CT257	Implantable cardiac pacemakers			A07	Electrical /Electronic Property Problem	14/06/2012	
									A0708	Power Problem	14/06/2012	
2026	2011	All other reportable incidents			CT257	Implantable cardiac pacemakers	F1905	Device Revision or Replacement	A07	Electrical /Electronic Property Problem	12/09/2011	
									A0708	Power Problem	12/09/2011	
2027	2011				CT257	Implantable cardiac pacemakers	F1905	Device Revision or Replacement	A05	Mechanical Problem		
									A0501	Detachment of Device or Device Component		
2028	2012				CT257	Implantable cardiac pacemakers	F21	Unexpected Deterioration	A26	Insufficient Device Problem Information		
2029	2012				CT257	Implantable cardiac pacemakers	F21	Unexpected Deterioration	A26	Insufficient Device Problem Information		
2030	2011	All other reportable incidents			CT257	Implantable cardiac pacemakers	F1905	Device Revision or Replacement	A07	Electrical /Electronic Property Problem		
									A0708	Power Problem		
2031	2012		47270	Cardiac resynchronization therapy implantable defibrillator	CT2866	Implantable cardiac electrical stimulators			A26	Insufficient Device Problem Information		
2032	2012		47270	Cardiac resynchronization therapy implantable defibrillator	CT2866	Implantable cardiac electrical stimulators			A26	Insufficient Device Problem Information		
2033	2012		47270	Cardiac resynchronization therapy implantable defibrillator	CT2866	Implantable cardiac electrical stimulators			A26	Insufficient Device Problem Information		
2034	2011		47270	Cardiac resynchronization therapy implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	F14	Prolonged Episode of Care	A26	Insufficient Device Problem Information		
2035	2012		47270	Cardiac resynchronization therapy implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E2401	Insufficient Information	A26	Insufficient Device Problem Information		
2036	2011		47270	Cardiac resynchronization therapy implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E2401	Insufficient Information	A26	Insufficient Device Problem Information		
2037	2011	All other reportable incidents	47270	Cardiac resynchronization therapy implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	F1905	Device Revision or Replacement	A07	Electrical /Electronic Property Problem	17/04/2011	
									A0708	Power Problem	17/04/2011	
2038	2011				CT257	Implantable cardiac pacemakers			A19	Environmental Compatibility Problem		
									A191002	Loss of Power		
2039	2012		47270	Cardiac resynchronization therapy implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E2401	Insufficient Information	A26	Insufficient Device Problem Information		
2040	2012		47270	Cardiac resynchronization therapy implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E2401	Insufficient Information	A26	Insufficient Device Problem Information		
2041	2012		47270	Cardiac resynchronization therapy implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E2401	Insufficient Information	A26	Insufficient Device Problem Information		

2042	2012				CT257	Implantable cardiac pacemakers	F15	Recognized Device or Procedural Complication	A26	Insufficient Device Problem Information		
2043	2012				CT257	Implantable cardiac pacemakers	F1905	Device Revision or Replacement	A07	Electrical /Electronic Property Problem		
									A0708	Power Problem		
2044	2011		47270	Cardiac resynchronization therapy implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E2401	Insufficient Information	A26	Insufficient Device Problem Information		
2045	2012	All other reportable incidents			CT257	Implantable cardiac pacemakers	F1905	Device Revision or Replacement	A13	Communication or Transmission Problem		
									A26	Insufficient Device Problem Information		
2046	2011	All other reportable incidents			CT257	Implantable cardiac pacemakers	F1905	Device Revision or Replacement	A07	Electrical /Electronic Property Problem	12/08/2011	
									A0708	Power Problem	12/08/2011	
2047	2012		47270	Cardiac resynchronization therapy implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	F15	Recognized Device or Procedural Complication	A26	Insufficient Device Problem Information		
2048	2012		47270	Cardiac resynchronization therapy implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E2401	Insufficient Information	A13	Communication or Transmission Problem		
									A26	Insufficient Device Problem Information		
2049	2011				CT257	Implantable cardiac pacemakers	F1905	Device Revision or Replacement	A1304	Telemetry Discrepancy		
									A26	Insufficient Device Problem Information		
2050	2012		47270	Cardiac resynchronization therapy implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	F21	Unexpected Deterioration	A26	Insufficient Device Problem Information		
2051	2012				CT257	Implantable cardiac pacemakers	F1905	Device Revision or Replacement	A07	Electrical /Electronic Property Problem		
									A0708	Power Problem		
2052	2012		47270	Cardiac resynchronization therapy implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	F1905	Device Revision or Replacement	A26	Insufficient Device Problem Information		
2053	2012				CT257	Implantable cardiac pacemakers	F1905	Device Revision or Replacement	A07	Electrical /Electronic Property Problem		
									A0708	Power Problem		
2054	2012		47270	Cardiac resynchronization therapy implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	F15	Recognized Device or Procedural Complication	A26	Insufficient Device Problem Information		
2055	2012		47270	Cardiac resynchronization therapy implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	F15	Recognized Device or Procedural Complication	A26	Insufficient Device Problem Information		
2056	2011				CT257	Implantable cardiac pacemakers	F14	Prolonged Episode of Care	A1104	Computer Operating System Problem		
2057	2011		47270	Cardiac resynchronization therapy implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	F1905	Device Revision or Replacement	A26	Insufficient Device Problem Information		
2058	2012		47270	Cardiac resynchronization therapy implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E2401	Insufficient Information	A13	Communication or Transmission Problem		
									A26	Insufficient Device Problem Information		
2059	2011	Death			CT257	Implantable cardiac pacemakers			A26	Insufficient Device Problem Information	04/01/2011	
2060	2012		47270	Cardiac resynchronization therapy implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	F1905	Device Revision or Replacement	A07	Electrical /Electronic Property Problem		
									A0709	Device Sensing Problem		
									A15	Activation, Positioning or Separation Problem		
									A150203	Difficult or Delayed Positioning		
2061	2012		47270	Cardiac resynchronization therapy implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	F1905	Device Revision or Replacement	A07	Electrical /Electronic Property Problem		
									A0709	Device Sensing Problem		
2062	2012				CT257	Implantable cardiac pacemakers	F1905	Device Revision or Replacement	A07	Electrical /Electronic Property Problem		
									A0721	Circuit Failure		
2063	2012	All other reportable incidents			CT257	Implantable cardiac pacemakers	F1905	Device Revision or Replacement	A07	Electrical /Electronic Property Problem	10/02/2012	
									A0709	Device Sensing Problem	10/02/2012	
2064	2012				CT257	Implantable cardiac pacemakers	E2401	Insufficient Information	A26	Insufficient Device Problem Information		
2065	2011				CT257	Implantable cardiac pacemakers	F1905	Device Revision or Replacement	A07	Electrical /Electronic Property Problem		
									A0709	Device Sensing Problem		
2066	2011				CT257	Implantable cardiac pacemakers	F1905	Device Revision or Replacement	A07	Electrical /Electronic Property Problem		
									A0708	Power Problem		
2067	2012		47270	Cardiac resynchronization therapy implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	F21	Unexpected Deterioration	A26	Insufficient Device Problem Information		
2068	2012		47270	Cardiac resynchronization therapy implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	F21	Unexpected Deterioration	A26	Insufficient Device Problem Information		
2069	2012		47270	Cardiac resynchronization therapy implantable defibrillator	CT2866	Implantable cardiac electrical stimulators			A26	Insufficient Device Problem Information		
2070	2012		47270	Cardiac resynchronization therapy implantable defibrillator	CT2866	Implantable cardiac electrical stimulators			A26	Insufficient Device Problem Information		
2071	2012		47270	Cardiac resynchronization therapy implantable defibrillator	CT2866	Implantable cardiac electrical stimulators			A26	Insufficient Device Problem Information		
2072	2012		47270	Cardiac resynchronization therapy implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	F15	Recognized Device or Procedural Complication	A07	Electrical /Electronic Property Problem		
									A0721	Circuit Failure		
2073	2012		47270	Cardiac resynchronization therapy implantable defibrillator	CT2866	Implantable cardiac electrical stimulators			A26	Insufficient Device Problem Information		
2074	2012		47270	Cardiac resynchronization therapy implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	F21	Unexpected Deterioration	A26	Insufficient Device Problem Information		
2075	2012		47270	Cardiac resynchronization therapy implantable defibrillator	CT2866	Implantable cardiac electrical stimulators			A26	Insufficient Device Problem Information		
2076	2012		47270	Cardiac resynchronization therapy implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	F21	Unexpected Deterioration	A26	Insufficient Device Problem Information		
2077	2012		47270	Cardiac resynchronization therapy implantable defibrillator	CT2866	Implantable cardiac electrical stimulators			A26	Insufficient Device Problem Information		
2078	2012		47270	Cardiac resynchronization therapy implantable defibrillator	CT2866	Implantable cardiac electrical stimulators			A26	Insufficient Device Problem Information		
2079	2012		47270	Cardiac resynchronization therapy implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	F15	Recognized Device or Procedural Complication	A26	Insufficient Device Problem Information		
2080	2012		47270	Cardiac resynchronization therapy implantable defibrillator	CT2866	Implantable cardiac electrical stimulators			A26	Insufficient Device Problem Information		
2081	2012		47270	Cardiac resynchronization therapy implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	F21	Unexpected Deterioration	A26	Insufficient Device Problem Information		
2082	2012		47270	Cardiac resynchronization therapy implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	F14	Prolonged Episode of Care	A26	Insufficient Device Problem Information		
2083	2012		35223	Endocardial/interventricular septal pacing lead	CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A23	Use of Device Problem		
2084	2011		47270	Cardiac resynchronization therapy implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E2401	Insufficient Information	A09	Output Problem		
									A090404	Output Above Specifications		
2085	2011		47270	Cardiac resynchronization therapy implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	F1905	Device Revision or Replacement	A09	Output Problem		
									A0907	No Device Output		
									A1304	Telemetry Discrepancy		
									A26	Insufficient Device Problem Information		
2086	2012				CT257	Implantable cardiac pacemakers	F1905	Device Revision or Replacement	A13	Communication or Transmission Problem		
									A26	Insufficient Device Problem Information		
2087	2012		47270	Cardiac resynchronization therapy implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	F1905	Device Revision or Replacement	A07	Electrical /Electronic Property Problem		
									A0708	Power Problem		
2088	2012		47270	Cardiac resynchronization therapy implantable	CT2866	Implantable cardiac electrical stimulators	E2401	Insufficient Information	A13	Communication or Transmission Problem		

				defibrillator					A26	Insufficient Device Problem Information		
2089	2012				CT257	Implantable cardiac pacemakers	F1905	Device Revision or Replacement	A09	Output Problem		
									A0907	No Device Output		
2090	2012		47270	Cardiac resynchronization therapy implantable defibrillator	CT2866	Implantable cardiac electrical stimulators			A07	Electrical /Electronic Property Problem		
									A0708	Power Problem		
2091	2012		47270	Cardiac resynchronization therapy implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	F15	Recognized Device or Procedural Complication	A26	Insufficient Device Problem Information		
2092	2011				CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A07	Electrical /Electronic Property Problem		
									A0708	Power Problem		
2093	2011		47270	Cardiac resynchronization therapy implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	F1905	Device Revision or Replacement	A07	Electrical /Electronic Property Problem		
									A0709	Device Sensing Problem		
2094	2012	TBD			CT257	Implantable cardiac pacemakers	E2401	Insufficient Information	A07	Electrical /Electronic Property Problem		
									A0709	Device Sensing Problem		
2095	2011				CT257	Implantable cardiac pacemakers	F1905	Device Revision or Replacement	A26	Insufficient Device Problem Information		
2096	2012		47270	Cardiac resynchronization therapy implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	F15	Recognized Device or Procedural Complication	A26	Insufficient Device Problem Information		
2097	2012				CT257	Implantable cardiac pacemakers			A07	Electrical /Electronic Property Problem		
									A0709	Device Sensing Problem		
2098	2011		35223	Endocardial/interventricular septal pacing lead	CT1684	Cardiac pacemakers and associated devices	F14	Prolonged Episode of Care	A04	Material Integrity Problem		
									A0405	Degraded		
2099	2012				CT257	Implantable cardiac pacemakers	F1905	Device Revision or Replacement	A1304	Telemetry Discrepancy		
									A26	Insufficient Device Problem Information		
2100	2012		47270	Cardiac resynchronization therapy implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	F15	Recognized Device or Procedural Complication	A26	Insufficient Device Problem Information		
2101	2011		47270	Cardiac resynchronization therapy implantable defibrillator	CT2866	Implantable cardiac electrical stimulators			A26	Insufficient Device Problem Information		
2102	2011		47270	Cardiac resynchronization therapy implantable defibrillator	CT2866	Implantable cardiac electrical stimulators			A26	Insufficient Device Problem Information		
2103	2012		47270	Cardiac resynchronization therapy implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	F15	Recognized Device or Procedural Complication	A26	Insufficient Device Problem Information		
2104	2012				CT257	Implantable cardiac pacemakers	F15	Recognized Device or Procedural Complication	A26	Insufficient Device Problem Information		
2105	2012	All other reportable incidents			CT257	Implantable cardiac pacemakers	E0402	Hypersensitivity/Allergic Reaction	A01	Patient Device Interaction Problem		
									A10	Temperature Problem		
									A1005	Overheating of Device		
									A17	Compatibility Problem		
							E1704	Burn(s)	A01	Patient Device Interaction Problem		
									A10	Temperature Problem		
									A1005	Overheating of Device		
									A17	Compatibility Problem		
2106	2011				CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A07	Electrical /Electronic Property Problem		
									A0708	Power Problem		
2107	2011		47270	Cardiac resynchronization therapy implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A07	Electrical /Electronic Property Problem		
									A0709	Device Sensing Problem		
2108	2011		47270	Cardiac resynchronization therapy implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E2109	Implant Pain	A26	Insufficient Device Problem Information		
2109	2011		47270	Cardiac resynchronization therapy implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	F1905	Device Revision or Replacement	A07	Electrical /Electronic Property Problem		
									A0709	Device Sensing Problem		
2110	2011	All other reportable incidents			CT257	Implantable cardiac pacemakers	F1905	Device Revision or Replacement	A07	Electrical /Electronic Property Problem	26/09/2011	
									A0708	Power Problem	26/09/2011	
2111	2012		47270	Cardiac resynchronization therapy implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	F15	Recognized Device or Procedural Complication	A26	Insufficient Device Problem Information		
2112	2011	All other reportable incidents			CT257	Implantable cardiac pacemakers	F1905	Device Revision or Replacement	A07	Electrical /Electronic Property Problem		
									A0708	Power Problem		
2113	2012	All other reportable incidents			CT257	Implantable cardiac pacemakers	F1905	Device Revision or Replacement	A07	Electrical /Electronic Property Problem	22/06/2012	
									A0708	Power Problem	22/06/2012	
2114	2012	All other reportable incidents			CT257	Implantable cardiac pacemakers	F1905	Device Revision or Replacement	A07	Electrical /Electronic Property Problem	09/11/2012	
									A0708	Power Problem	09/11/2012	
2115	2011				CT257	Implantable cardiac pacemakers	F1905	Device Revision or Replacement	A13	Communication or Transmission Problem		
									A26	Insufficient Device Problem Information		
2116	2012				CT257	Implantable cardiac pacemakers	F15	Recognized Device or Procedural Complication	A26	Insufficient Device Problem Information		
2117	2012		47270	Cardiac resynchronization therapy implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	F21	Unexpected Deterioration	A26	Insufficient Device Problem Information		
2118	2012		47270	Cardiac resynchronization therapy implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	F21	Unexpected Deterioration	A26	Insufficient Device Problem Information		
2119	2012		47270	Cardiac resynchronization therapy implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	F21	Unexpected Deterioration	A26	Insufficient Device Problem Information		
2120	2012		47270	Cardiac resynchronization therapy implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	F21	Unexpected Deterioration	A26	Insufficient Device Problem Information		
2121	2012		47270	Cardiac resynchronization therapy implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E2401	Insufficient Information	A26	Insufficient Device Problem Information		
2122	2012		47270	Cardiac resynchronization therapy implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E1906	Unspecified Infection	A26	Insufficient Device Problem Information		
2123	2012		47270	Cardiac resynchronization therapy implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	F21	Unexpected Deterioration	A26	Insufficient Device Problem Information		
2124	2012		47270	Cardiac resynchronization therapy implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	F21	Unexpected Deterioration	A26	Insufficient Device Problem Information		
2125	2011		47270	Cardiac resynchronization therapy implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E2401	Insufficient Information	A07	Electrical /Electronic Property Problem		
									A0709	Device Sensing Problem		
2126	2012		47270	Cardiac resynchronization therapy implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	F21	Unexpected Deterioration	A26	Insufficient Device Problem Information		
2127	2012		35223	Endocardial/interventricular septal pacing lead	CT1684	Cardiac pacemakers and associated devices	F1905	Device Revision or Replacement	A15	Activation, Positioning or Separation Problem		
									A150301	Separation Failure		
2128	2012		47270	Cardiac resynchronization therapy implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E2401	Insufficient Information	A26	Insufficient Device Problem Information		
2129	2012		47270	Cardiac resynchronization therapy implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	F14	Prolonged Episode of Care	A26	Insufficient Device Problem Information		
2130	2012		47270	Cardiac resynchronization therapy implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E2401	Insufficient Information	A26	Insufficient Device Problem Information		
2131	2012		47270	Cardiac resynchronization therapy implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	F14	Prolonged Episode of Care	A26	Insufficient Device Problem Information		
2132	2012				CT257	Implantable cardiac pacemakers	F1905	Device Revision or Replacement	A07	Electrical /Electronic Property Problem		
									A0708	Power Problem		
2133	2011				CT257	Implantable cardiac pacemakers	E2401	Insufficient Information	A1104	Computer Operating System Problem		

2134	2011				CT257	Implantable cardiac pacemakers	F1905	Device Revision or Replacement	A13	Communication or Transmission Problem		
									A26	Insufficient Device Problem Information		
2135	2011				CT257	Implantable cardiac pacemakers	F1905	Device Revision or Replacement	A13	Communication or Transmission Problem		
									A26	Insufficient Device Problem Information		
2136	2011		47270	Cardiac resynchronization therapy implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E2401	Insufficient Information	A26	Insufficient Device Problem Information		
2137	2011				CT257	Implantable cardiac pacemakers	F1905	Device Revision or Replacement	A07	Electrical /Electronic Property Problem		
									A0709	Device Sensing Problem		
2138	2012	All other reportable incidents			CT257	Implantable cardiac pacemakers	F1905	Device Revision or Replacement	A07	Electrical /Electronic Property Problem	16/11/2012	
									A0708	Power Problem	16/11/2012	
2139	2011				CT257	Implantable cardiac pacemakers	E2401	Insufficient Information	A07	Electrical /Electronic Property Problem		
									A0709	Device Sensing Problem		
2140	2011				CT257	Implantable cardiac pacemakers	E2401	Insufficient Information	A07	Electrical /Electronic Property Problem		
									A0709	Device Sensing Problem		
2141	2011				CT257	Implantable cardiac pacemakers	E2401	Insufficient Information	A07	Electrical /Electronic Property Problem		
									A0709	Device Sensing Problem		
2142	2011	All other reportable incidents			CT257	Implantable cardiac pacemakers	F1905	Device Revision or Replacement	A07	Electrical /Electronic Property Problem		
									A0708	Power Problem		
2143	2012				CT257	Implantable cardiac pacemakers			A07	Electrical /Electronic Property Problem		
									A0708	Power Problem		
2144	2011				CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A07	Electrical /Electronic Property Problem		
									A0709	Device Sensing Problem		
2145	2012				CT257	Implantable cardiac pacemakers	F1905	Device Revision or Replacement	A07	Electrical /Electronic Property Problem		
									A0708	Power Problem		
2146	2011				CT257	Implantable cardiac pacemakers	F1905	Device Revision or Replacement	A15	Activation, Positioning or Separation Problem	26/05/2011	
									A150101	Activation Failure	26/05/2011	
2147	2011				CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A11	Computer Software Problem		
									A1102	Application Program Problem		
2148	2012		47270	Cardiac resynchronization therapy implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A26	Insufficient Device Problem Information		
2149	2012				CT257	Implantable cardiac pacemakers	F1905	Device Revision or Replacement	A07	Electrical /Electronic Property Problem		
									A0708	Power Problem		
2150	2011				CT257	Implantable cardiac pacemakers			A07	Electrical /Electronic Property Problem		
									A0708	Power Problem		
2151	2011	All other reportable incidents	47270	Cardiac resynchronization therapy implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	F1905	Device Revision or Replacement	A07	Electrical /Electronic Property Problem	11/08/2011	
									A0709	Device Sensing Problem	11/08/2011	
2152	2011		35223	Endocardial/interventricular septal pacing lead	CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A04	Material Integrity Problem		
									A0404	Crack		
2153	2011	Death	47270	Cardiac resynchronization therapy implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E2401	Insufficient Information	A26	Insufficient Device Problem Information		
2154	2012		47270	Cardiac resynchronization therapy implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	F1905	Device Revision or Replacement	A07	Electrical /Electronic Property Problem		
									A0708	Power Problem		
2155	2011	Death			CT1684	Cardiac pacemakers and associated devices	E2401	Insufficient Information	A15	Activation, Positioning or Separation Problem		
									A150101	Activation Failure		
2156	2012		47270	Cardiac resynchronization therapy implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A26	Insufficient Device Problem Information		
2157	2011		47270	Cardiac resynchronization therapy implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	F1905	Device Revision or Replacement	A07	Electrical /Electronic Property Problem		
									A0709	Device Sensing Problem		
2158	2011				CT257	Implantable cardiac pacemakers	F1905	Device Revision or Replacement	A07	Electrical /Electronic Property Problem		
									A0708	Power Problem		
2159	2011				CT257	Implantable cardiac pacemakers	F1905	Device Revision or Replacement	A07	Electrical /Electronic Property Problem		
									A0708	Power Problem		
2160	2011				CT257	Implantable cardiac pacemakers	F1905	Device Revision or Replacement	A15	Activation, Positioning or Separation Problem		
									A150101	Activation Failure		
2161	2012				CT257	Implantable cardiac pacemakers	F15	Recognized Device or Procedural Complication	A26	Insufficient Device Problem Information		
2162	2011	All other reportable incidents			CT257	Implantable cardiac pacemakers	F1905	Device Revision or Replacement	A19	Environmental Compatibility Problem		
									A191002	Loss of Power		
2163	2012				CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A13	Communication or Transmission Problem		
									A26	Insufficient Device Problem Information		
2164	2012	All other reportable incidents			CT257	Implantable cardiac pacemakers	F1905	Device Revision or Replacement	A07	Electrical /Electronic Property Problem	05/03/2012	
									A0708	Power Problem	05/03/2012	
2165	2011				CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A07	Electrical /Electronic Property Problem		
									A0708	Power Problem		
2166	2011				CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A07	Electrical /Electronic Property Problem		
									A0708	Power Problem		
2167	2011				CT257	Implantable cardiac pacemakers			A12	Connection Problem		
2168	2012	TBD			CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A09	Output Problem		
									A0907	No Device Output		
2169	2011	All other reportable incidents	35223	Endocardial/interventricular septal pacing lead	CT1684	Cardiac pacemakers and associated devices	E2401	Insufficient Information	A07	Electrical /Electronic Property Problem		
									A0709	Device Sensing Problem		
2170	2011	All other reportable incidents			CT257	Implantable cardiac pacemakers	F1905	Device Revision or Replacement	A07	Electrical /Electronic Property Problem	14/10/2011	
									A0708	Power Problem	14/10/2011	
2171	2011	Death			CT1684	Cardiac pacemakers and associated devices	E0752	Unspecified Respiratory Problem	A07	Electrical /Electronic Property Problem		
									A0708	Power Problem		
2172	2012	All other reportable incidents	47270	Cardiac resynchronization therapy implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	F1905	Device Revision or Replacement	A07	Electrical /Electronic Property Problem	09/11/2012	
									A0708	Power Problem	09/11/2012	
2173	2012	All other reportable incidents			CT1684	Cardiac pacemakers and associated devices			A16	Protective Measures Problem		
									A1602	Fail-Safe Problem		
2174	2012				CT257	Implantable cardiac pacemakers	F1905	Device Revision or Replacement	A07	Electrical /Electronic Property Problem		
									A0708	Power Problem		
2175	2012	All other reportable incidents			CT257	Implantable cardiac pacemakers	F1905	Device Revision or Replacement	A07	Electrical /Electronic Property Problem		
									A0708	Power Problem		
2176	2012	All other reportable incidents			CT257	Implantable cardiac pacemakers	F1905	Device Revision or Replacement	A07	Electrical /Electronic Property Problem		
									A0708	Power Problem		
2177	2012		47270	Cardiac resynchronization therapy implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	F1905	Device Revision or Replacement	A07	Electrical /Electronic Property Problem		
									A0708	Power Problem		
2178	2011	All other reportable incidents			CT257	Implantable cardiac pacemakers	F1905	Device Revision or Replacement	A07	Electrical /Electronic Property Problem		
									A0708	Power Problem		

2179	2011		47270	Cardiac resynchronization therapy implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	F14	Prolonged Episode of Care	A26	Insufficient Device Problem Information		
2180	2012		47270	Cardiac resynchronization therapy implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E2401	Insufficient Information	A26	Insufficient Device Problem Information		
2181	2011				CT257	Implantable cardiac pacemakers			A26	Insufficient Device Problem Information		
2182	2012				CT257	Implantable cardiac pacemakers	F15	Recognized Device or Procedural Complication	A26	Insufficient Device Problem Information		
2183	2012		47270	Cardiac resynchronization therapy implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E2401	Insufficient Information	A26	Insufficient Device Problem Information		
2184	2012		47270	Cardiac resynchronization therapy implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	F15	Recognized Device or Procedural Complication	A26	Insufficient Device Problem Information		
2185	2012		47270	Cardiac resynchronization therapy implantable defibrillator	CT2866	Implantable cardiac electrical stimulators			A26	Insufficient Device Problem Information		
2186	2012		47270	Cardiac resynchronization therapy implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E2401	Insufficient Information	A26	Insufficient Device Problem Information		
2187	2012		47270	Cardiac resynchronization therapy implantable defibrillator	CT2866	Implantable cardiac electrical stimulators			A26	Insufficient Device Problem Information		
2188	2012		47270	Cardiac resynchronization therapy implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E1906	Unspecified Infection	A26	Insufficient Device Problem Information		
2189	2012		47270	Cardiac resynchronization therapy implantable defibrillator	CT2866	Implantable cardiac electrical stimulators			A26	Insufficient Device Problem Information		
2190	2012		47270	Cardiac resynchronization therapy implantable defibrillator	CT2866	Implantable cardiac electrical stimulators			A26	Insufficient Device Problem Information		
2191	2012		47270	Cardiac resynchronization therapy implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E2401	Insufficient Information	A26	Insufficient Device Problem Information		
2192	2012		47270	Cardiac resynchronization therapy implantable defibrillator	CT2866	Implantable cardiac electrical stimulators			A26	Insufficient Device Problem Information		
2193	2012		47270	Cardiac resynchronization therapy implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E2401	Insufficient Information	A26	Insufficient Device Problem Information		
2194	2012		47270	Cardiac resynchronization therapy implantable defibrillator	CT2866	Implantable cardiac electrical stimulators			A26	Insufficient Device Problem Information		
2195	2012		47270	Cardiac resynchronization therapy implantable defibrillator	CT2866	Implantable cardiac electrical stimulators			A26	Insufficient Device Problem Information		
2196	2012		47270	Cardiac resynchronization therapy implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E2401	Insufficient Information	A26	Insufficient Device Problem Information		
2197	2012		47270	Cardiac resynchronization therapy implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E2401	Insufficient Information	A26	Insufficient Device Problem Information		
2198	2012		47270	Cardiac resynchronization therapy implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	F14	Prolonged Episode of Care	A26	Insufficient Device Problem Information		
2199	2012				CT257	Implantable cardiac pacemakers			A07	Electrical /Electronic Property Problem		
									A0708	Power Problem		
									A09	Output Problem		
									A0907	No Device Output		
2200	2012		47270	Cardiac resynchronization therapy implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E1906	Unspecified Infection	A26	Insufficient Device Problem Information		
2201	2012		47270	Cardiac resynchronization therapy implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	F15	Recognized Device or Procedural Complication	A26	Insufficient Device Problem Information		
2202	2012		47270	Cardiac resynchronization therapy implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	F21	Unexpected Deterioration	A26	Insufficient Device Problem Information		
2203	2012		47270	Cardiac resynchronization therapy implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	F21	Unexpected Deterioration	A26	Insufficient Device Problem Information		
2204	2011				CT257	Implantable cardiac pacemakers			A07	Electrical /Electronic Property Problem		
									A0708	Power Problem		
2205	2012				CT257	Implantable cardiac pacemakers	F1905	Device Revision or Replacement	A07	Electrical /Electronic Property Problem		
									A0709	Device Sensing Problem		
2206	2012				CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A26	Insufficient Device Problem Information		
2207	2012				CT257	Implantable cardiac pacemakers	F15	Recognized Device or Procedural Complication	A26	Insufficient Device Problem Information		
2208	2011				CT257	Implantable cardiac pacemakers	F1905	Device Revision or Replacement	A1304	Telemetry Discrepancy		
									A26	Insufficient Device Problem Information		
2209	2012	All other reportable incidents	47270	Cardiac resynchronization therapy implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	F1905	Device Revision or Replacement	A07	Electrical /Electronic Property Problem		
									A0708	Power Problem		
2210	2012	All other reportable incidents			CT257	Implantable cardiac pacemakers	E2401	Insufficient Information	A13	Communication or Transmission Problem		
									A26	Insufficient Device Problem Information		
2211	2012	All other reportable incidents			CT257	Implantable cardiac pacemakers	E0402	Hypersensitivity/Allergic Reaction	A07	Electrical /Electronic Property Problem	03/07/2012	
									A0708	Power Problem	03/07/2012	
									A09	Output Problem	03/07/2012	
									A0908	Incorrect, Inadequate or Imprecise Result or Readings	03/07/2012	
							F1905	Device Revision or Replacement	A07	Electrical /Electronic Property Problem	03/07/2012	
									A0708	Power Problem	03/07/2012	
									A09	Output Problem	03/07/2012	
									A0908	Incorrect, Inadequate or Imprecise Result or Readings	03/07/2012	
2212	2012		47270	Cardiac resynchronization therapy implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	F21	Unexpected Deterioration	A26	Insufficient Device Problem Information		
2213	2011	All other reportable incidents			CT257	Implantable cardiac pacemakers	F1905	Device Revision or Replacement	A07	Electrical /Electronic Property Problem		
									A0708	Power Problem		
2214	2011	All other reportable incidents			CT257	Implantable cardiac pacemakers	F1905	Device Revision or Replacement	A07	Electrical /Electronic Property Problem	11/10/2011	
									A0708	Power Problem	11/10/2011	
2215	2011		47270	Cardiac resynchronization therapy implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A26	Insufficient Device Problem Information		
2216	2012				CT257	Implantable cardiac pacemakers	F1905	Device Revision or Replacement	A13	Communication or Transmission Problem		
									A26	Insufficient Device Problem Information		
2217	2012	All other reportable incidents	47270	Cardiac resynchronization therapy implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	F1905	Device Revision or Replacement	A07	Electrical /Electronic Property Problem		
									A0708	Power Problem		
2218	2012	All other reportable incidents			CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A09	Output Problem		
									A0907	No Device Output		
2219	2012	All other reportable incidents			CT257	Implantable cardiac pacemakers	F1905	Device Revision or Replacement	A07	Electrical /Electronic Property Problem	13/11/2012	
									A0708	Power Problem	13/11/2012	
2220	2012	All other reportable incidents			CT257	Implantable cardiac pacemakers	F1905	Device Revision or Replacement	A07	Electrical /Electronic Property Problem	30/04/2012	
									A0708	Power Problem	30/04/2012	
									A13	Communication or Transmission Problem	30/04/2012	
									A26	Insufficient Device Problem Information	30/04/2012	
2221	2012		47270	Cardiac resynchronization therapy implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	F14	Prolonged Episode of Care	A26	Insufficient Device Problem Information		

2222	2012		47270	Cardiac resynchronization therapy implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	F15	Recognized Device or Procedural Complication	A26	Insufficient Device Problem Information		
2223	2011	All other reportable incidents			CT257	Implantable cardiac pacemakers	F1905	Device Revision or Replacement	A07	Electrical /Electronic Property Problem		
									A0708	Power Problem		
2224	2012				CT257	Implantable cardiac pacemakers	E0514	Thrombosis/Thrombus	A26	Insufficient Device Problem Information		
2225	2011		35223	Endocardial/interventricular septal pacing lead	CT1684	Cardiac pacemakers and associated devices	F1905	Device Revision or Replacement	A07	Electrical /Electronic Property Problem		
									A0709	Device Sensing Problem		
2226	2012	All other reportable incidents	47270	Cardiac resynchronization therapy implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	F1905	Device Revision or Replacement	A07	Electrical /Electronic Property Problem	11/07/2012	
									A0708	Power Problem	11/07/2012	
2227	2012				CT257	Implantable cardiac pacemakers	F15	Recognized Device or Procedural Complication	A26	Insufficient Device Problem Information		
2228	2011		47270	Cardiac resynchronization therapy implantable defibrillator	CT2866	Implantable cardiac electrical stimulators			A26	Insufficient Device Problem Information		
2229	2012				CT257	Implantable cardiac pacemakers	F15	Recognized Device or Procedural Complication	A26	Insufficient Device Problem Information		
2230	2012		47270	Cardiac resynchronization therapy implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E1603	Bone Fracture(s)	A26	Insufficient Device Problem Information		
2231	2012		47270	Cardiac resynchronization therapy implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	F14	Prolonged Episode of Care	A26	Insufficient Device Problem Information		
2232	2012		47270	Cardiac resynchronization therapy implantable defibrillator	CT2866	Implantable cardiac electrical stimulators			A26	Insufficient Device Problem Information		
2233	2012		47270	Cardiac resynchronization therapy implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E2401	Insufficient Information	A26	Insufficient Device Problem Information		
2234	2012		47270	Cardiac resynchronization therapy implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E2401	Insufficient Information	A26	Insufficient Device Problem Information		
2235	2011	All other reportable incidents			CT257	Implantable cardiac pacemakers	F1905	Device Revision or Replacement	A07	Electrical /Electronic Property Problem		
									A0708	Power Problem		
2236	2012		47270	Cardiac resynchronization therapy implantable defibrillator	CT2866	Implantable cardiac electrical stimulators			A26	Insufficient Device Problem Information		
2237	2012		47270	Cardiac resynchronization therapy implantable defibrillator	CT2866	Implantable cardiac electrical stimulators			A26	Insufficient Device Problem Information		
2238	2012		47270	Cardiac resynchronization therapy implantable defibrillator	CT2866	Implantable cardiac electrical stimulators			A26	Insufficient Device Problem Information		
2239	2012		47270	Cardiac resynchronization therapy implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E2401	Insufficient Information	A26	Insufficient Device Problem Information		
2240	2012		47270	Cardiac resynchronization therapy implantable defibrillator	CT2866	Implantable cardiac electrical stimulators			A26	Insufficient Device Problem Information		
2241	2012		47270	Cardiac resynchronization therapy implantable defibrillator	CT2866	Implantable cardiac electrical stimulators			A26	Insufficient Device Problem Information		
2242	2012		47270	Cardiac resynchronization therapy implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E2401	Insufficient Information	A26	Insufficient Device Problem Information		
2243	2012		47270	Cardiac resynchronization therapy implantable defibrillator	CT2866	Implantable cardiac electrical stimulators			A26	Insufficient Device Problem Information		
2244	2012		47270	Cardiac resynchronization therapy implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E2401	Insufficient Information	A26	Insufficient Device Problem Information		
2245	2012		47270	Cardiac resynchronization therapy implantable defibrillator	CT2866	Implantable cardiac electrical stimulators			A26	Insufficient Device Problem Information		
2246	2012		47263	Cardiac resynchronization therapy implantable pacemaker	CT1684	Cardiac pacemakers and associated devices	E2401	Insufficient Information	A26	Insufficient Device Problem Information		
					CT1685	Cardiac pacemakers	E2401	Insufficient Information	A26	Insufficient Device Problem Information		
					CT257	Implantable cardiac pacemakers	E2401	Insufficient Information	A26	Insufficient Device Problem Information		
					CT2866	Implantable cardiac electrical stimulators	E2401	Insufficient Information	A26	Insufficient Device Problem Information		
2247	2012	All other reportable incidents			CT257	Implantable cardiac pacemakers	F1905	Device Revision or Replacement	A09	Output Problem	24/07/2012	
									A0907	No Device Output	24/07/2012	
2248	2012	All other reportable incidents			CT257	Implantable cardiac pacemakers	F1905	Device Revision or Replacement	A09	Output Problem	09/11/2011	
									A0907	No Device Output	09/11/2011	
2249	2012		47270	Cardiac resynchronization therapy implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	F1905	Device Revision or Replacement	A07	Electrical /Electronic Property Problem		
									A0708	Power Problem		
2250	2011	All other reportable incidents			CT257	Implantable cardiac pacemakers	F1905	Device Revision or Replacement	A09	Output Problem		
									A0908	Incorrect, Inadequate or Imprecise Result or Readings		
2251	2011		47270	Cardiac resynchronization therapy implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E0402	Hypersensitivity/Allergic Reaction	A26	Insufficient Device Problem Information		
2252	2011	All other reportable incidents			CT257	Implantable cardiac pacemakers	F1905	Device Revision or Replacement	A07	Electrical /Electronic Property Problem		
									A0708	Power Problem		
2253	2012	All other reportable incidents			CT257	Implantable cardiac pacemakers	F1905	Device Revision or Replacement	A07	Electrical /Electronic Property Problem		
									A0708	Power Problem		
									A09	Output Problem		
									A0907	No Device Output		
									A1104	Computer Operating System Problem		
2254	2012		35223	Endocardial/interventricular septal pacing lead	CT1684	Cardiac pacemakers and associated devices	E2401	Insufficient Information	A26	Insufficient Device Problem Information		
2255	2011		35223	Endocardial/interventricular septal pacing lead	CT1684	Cardiac pacemakers and associated devices	F1905	Device Revision or Replacement	A12	Connection Problem		
									A1203	Disconnection		
2256	2012				CT257	Implantable cardiac pacemakers	F1905	Device Revision or Replacement	A26	Insufficient Device Problem Information		
2257	2011				CT257	Implantable cardiac pacemakers	E2401	Insufficient Information	A26	Insufficient Device Problem Information		
2258	2012		47270	Cardiac resynchronization therapy implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	F14	Prolonged Episode of Care	A07	Electrical /Electronic Property Problem		
									A0708	Power Problem		
2259	2012				CT257	Implantable cardiac pacemakers	F1905	Device Revision or Replacement	A07	Electrical /Electronic Property Problem		
									A0708	Power Problem		
2260	2011		47270	Cardiac resynchronization therapy implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E2401	Insufficient Information	A26	Insufficient Device Problem Information		
2261	2012		47270	Cardiac resynchronization therapy implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	F1905	Device Revision or Replacement	A07	Electrical /Electronic Property Problem		
									A0708	Power Problem		
2262	2011	All other reportable incidents			CT1684	Cardiac pacemakers and associated devices	E2401	Insufficient Information	A09	Output Problem		
									A0907	No Device Output		
									A15	Activation, Positioning or Separation Problem		
									A150101	Activation Failure		
2263	2011		47270	Cardiac resynchronization therapy implantable defibrillator	CT2866	Implantable cardiac electrical stimulators			A26	Insufficient Device Problem Information		
2264	2011		47270	Cardiac resynchronization therapy implantable defibrillator	CT2866	Implantable cardiac electrical stimulators			A26	Insufficient Device Problem Information		
2265	2011				CT257	Implantable cardiac pacemakers	F1905	Device Revision or Replacement	A13	Communication or Transmission Problem		
									A26	Insufficient Device Problem Information		
2266	2012				CT257	Implantable cardiac pacemakers	F1905	Device Revision or Replacement	A07	Electrical /Electronic Property Problem		
									A0708	Power Problem		

2267	2011				CT257	Implantable cardiac pacemakers	E2401	Insufficient Information	A07	Electrical /Electronic Property Problem		
									A0708	Power Problem		
2268	2011				CT257	Implantable cardiac pacemakers	F1905	Device Revision or Replacement	A07	Electrical /Electronic Property Problem		
									A0708	Power Problem		
2269	2012				CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A07	Electrical /Electronic Property Problem		
									A0708	Power Problem		
2270	2012				CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A26	Insufficient Device Problem Information		
2271	2012		47270	Cardiac resynchronization therapy implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	F1905	Device Revision or Replacement	A07	Electrical /Electronic Property Problem		
									A0708	Power Problem		
2272	2012		47270	Cardiac resynchronization therapy implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E2104	Electric Shock	A09	Output Problem		
									A090404	Output Above Specifications		
2273	2012				CT257	Implantable cardiac pacemakers	E2401	Insufficient Information	A26	Insufficient Device Problem Information		
2274	2012				CT257	Implantable cardiac pacemakers	F15	Recognized Device or Procedural Complication	A26	Insufficient Device Problem Information		
2275	2012		47263	Cardiac resynchronization therapy implantable pacemaker	CT1684	Cardiac pacemakers and associated devices	F15	Recognized Device or Procedural Complication	A26	Insufficient Device Problem Information		
					CT1685	Cardiac pacemakers	F15	Recognized Device or Procedural Complication	A26	Insufficient Device Problem Information		
					CT257	Implantable cardiac pacemakers	F15	Recognized Device or Procedural Complication	A26	Insufficient Device Problem Information		
					CT2866	Implantable cardiac electrical stimulators	F15	Recognized Device or Procedural Complication	A26	Insufficient Device Problem Information		
2276	2012		47270	Cardiac resynchronization therapy implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	F15	Recognized Device or Procedural Complication	A26	Insufficient Device Problem Information		
2277	2011		47270	Cardiac resynchronization therapy implantable defibrillator	CT2866	Implantable cardiac electrical stimulators			A07	Electrical /Electronic Property Problem		
									A0709	Device Sensing Problem		
2278	2012		47270	Cardiac resynchronization therapy implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E2401	Insufficient Information	A26	Insufficient Device Problem Information		
2279	2011		47270	Cardiac resynchronization therapy implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	F1905	Device Revision or Replacement	A0104	Migration or Expulsion of Device		
2280	2012				CT257	Implantable cardiac pacemakers	F14	Prolonged Episode of Care	A15	Activation, Positioning or Separation Problem		
									A150203	Difficult or Delayed Positioning		
2281	2012				CT257	Implantable cardiac pacemakers	F1905	Device Revision or Replacement	A09	Output Problem		
									A090404	Output Above Specifications		
2282	2011	All other reportable incidents	47270	Cardiac resynchronization therapy implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E2401	Insufficient Information	A07	Electrical /Electronic Property Problem	23/08/2011	
									A0708	Power Problem	23/08/2011	
2283	2012				CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A17	Compatibility Problem		
									A1702	Device-Device Incompatibility		
2284	2012		47270	Cardiac resynchronization therapy implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	F1905	Device Revision or Replacement	A09	Output Problem		
									A090404	Output Above Specifications		
2285	2012		47270	Cardiac resynchronization therapy implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	F15	Recognized Device or Procedural Complication	A26	Insufficient Device Problem Information		
2286	2012		47270	Cardiac resynchronization therapy implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	F15	Recognized Device or Procedural Complication	A26	Insufficient Device Problem Information		
2287	2012		47270	Cardiac resynchronization therapy implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E2401	Insufficient Information	A26	Insufficient Device Problem Information		
2288	2012		47270	Cardiac resynchronization therapy implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E2401	Insufficient Information	A26	Insufficient Device Problem Information		
2289	2012		47270	Cardiac resynchronization therapy implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E2401	Insufficient Information	A26	Insufficient Device Problem Information		
2290	2012		47270	Cardiac resynchronization therapy implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E2401	Insufficient Information	A26	Insufficient Device Problem Information		
2291	2012		47270	Cardiac resynchronization therapy implantable defibrillator	CT2866	Implantable cardiac electrical stimulators			A26	Insufficient Device Problem Information		
2292	2012		47270	Cardiac resynchronization therapy implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E2401	Insufficient Information	A26	Insufficient Device Problem Information		
2293	2012		47270	Cardiac resynchronization therapy implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E2401	Insufficient Information	A26	Insufficient Device Problem Information		
2294	2012		47270	Cardiac resynchronization therapy implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	F21	Unexpected Deterioration	A26	Insufficient Device Problem Information		
2295	2012				CT257	Implantable cardiac pacemakers	F1905	Device Revision or Replacement	A07	Electrical /Electronic Property Problem		
									A0708	Power Problem		
2296	2012				CT257	Implantable cardiac pacemakers	E2401	Insufficient Information	A07	Electrical /Electronic Property Problem		
									A0709	Device Sensing Problem		
2297	2012		47270	Cardiac resynchronization therapy implantable defibrillator	CT2866	Implantable cardiac electrical stimulators			A09	Output Problem		
									A0908	Incorrect, Inadequate or Imprecise Result or Readings		
2298	2012		47270	Cardiac resynchronization therapy implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E2401	Insufficient Information	A26	Insufficient Device Problem Information		
2299	2012	All other reportable incidents			CT257	Implantable cardiac pacemakers	F1905	Device Revision or Replacement	A07	Electrical /Electronic Property Problem	05/07/2012	
									A0708	Power Problem	05/07/2012	
2300	2012		47270	Cardiac resynchronization therapy implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E2104	Electric Shock	A09	Output Problem		
									A090404	Output Above Specifications		
2301	2012	All other reportable incidents			CT257	Implantable cardiac pacemakers	F1905	Device Revision or Replacement	A07	Electrical /Electronic Property Problem		
									A0708	Power Problem		
2302	2012	All other reportable incidents			CT257	Implantable cardiac pacemakers	F1905	Device Revision or Replacement	A07	Electrical /Electronic Property Problem		
									A0708	Power Problem		
2303	2012		35039	Epicardial pacing lead	CT1684	Cardiac pacemakers and associated devices			A07	Electrical /Electronic Property Problem		
									A0709	Device Sensing Problem		
2304	2012				CT257	Implantable cardiac pacemakers	F1905	Device Revision or Replacement	A07	Electrical /Electronic Property Problem		
									A0708	Power Problem		
2305	2011	All other reportable incidents			CT257	Implantable cardiac pacemakers			A07	Electrical /Electronic Property Problem	25/07/2011	
									A0708	Power Problem	25/07/2011	
2306	2012		47270	Cardiac resynchronization therapy implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	F1905	Device Revision or Replacement	A07	Electrical /Electronic Property Problem		
									A0708	Power Problem		
2307	2012		35039	Epicardial pacing lead	CT1684	Cardiac pacemakers and associated devices	F1905	Device Revision or Replacement	A04	Material Integrity Problem		
									A0404	Crack		
2308	2011	All other reportable incidents			CT257	Implantable cardiac pacemakers	E2401	Insufficient Information	A13	Communication or Transmission Problem	27/10/2011	
									A26	Insufficient Device Problem Information	27/10/2011	
2309	2012				CT257	Implantable cardiac pacemakers	F1905	Device Revision or Replacement	A07	Electrical /Electronic Property Problem		
									A0708	Power Problem		
2310	2012		47270	Cardiac resynchronization therapy implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E2104	Electric Shock	A09	Output Problem		
									A090404	Output Above Specifications		
2311	2012		47270	Cardiac resynchronization therapy implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	F15	Recognized Device or Procedural Complication	A26	Insufficient Device Problem Information		
2312	2012		47270	Cardiac resynchronization therapy implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E2401	Insufficient Information	A26	Insufficient Device Problem Information		

2313	2012		47270	Cardiac resynchronization therapy implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E1603	Bone Fracture(s)	A26	Insufficient Device Problem Information		
2314	2013				CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A15	Activation, Positioning or Separation Problem		
									A150101	Activation Failure		
2315	2013				CT257	Implantable cardiac pacemakers	F1905	Device Revision or Replacement	A09	Output Problem		
									A0907	No Device Output		
									A1304	Telemetry Discrepancy		
									A26	Insufficient Device Problem Information		
2316	2013		47270	Cardiac resynchronization therapy implantable defibrillator	CT2866	Implantable cardiac electrical stimulators			A09	Output Problem		
									A0908	Incorrect, Inadequate or Imprecise Result or Readings		
2317	2013				CT257	Implantable cardiac pacemakers	F1905	Device Revision or Replacement	A07	Electrical /Electronic Property Problem		
									A0709	Device Sensing Problem		
2318	2014				CT257	Implantable cardiac pacemakers	E0402	Hypersensitivity/Allergic Reaction	A26	Insufficient Device Problem Information		
2319	2013		47270	Cardiac resynchronization therapy implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A07	Electrical /Electronic Property Problem		
									A0709	Device Sensing Problem		
2320	2013		47270	Cardiac resynchronization therapy implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	F1905	Device Revision or Replacement	A1104	Computer Operating System Problem		
2321	2014	All other reportable incidents			CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A09	Output Problem		
									A0908	Incorrect, Inadequate or Imprecise Result or Readings		
2322	2013				CT257	Implantable cardiac pacemakers	E2401	Insufficient Information	A07	Electrical /Electronic Property Problem		
									A0709	Device Sensing Problem		
2323	2014		47270	Cardiac resynchronization therapy implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E0402	Hypersensitivity/Allergic Reaction	A26	Insufficient Device Problem Information		
2324	2014		47270	Cardiac resynchronization therapy implantable defibrillator	CT2866	Implantable cardiac electrical stimulators			A26	Insufficient Device Problem Information		
2325	2014				CT257	Implantable cardiac pacemakers			A26	Insufficient Device Problem Information		
2326	2013				CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A07	Electrical /Electronic Property Problem		
									A0709	Device Sensing Problem		
2327	2013				CT257	Implantable cardiac pacemakers	F1905	Device Revision or Replacement	A07	Electrical /Electronic Property Problem		
									A0708	Power Problem		
							F21	Unexpected Deterioration	A07	Electrical /Electronic Property Problem		
									A0708	Power Problem		
2328	2014	All other reportable incidents			CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A07	Electrical /Electronic Property Problem		
									A0708	Power Problem		
2329	2013				CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A07	Electrical /Electronic Property Problem		
									A0708	Power Problem		
2330	2013				CT257	Implantable cardiac pacemakers	F1905	Device Revision or Replacement	A07	Electrical /Electronic Property Problem		
									A0708	Power Problem		
2331	2013		47270	Cardiac resynchronization therapy implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	F15	Recognized Device or Procedural Complication	A26	Insufficient Device Problem Information		
2332	2013				CT257	Implantable cardiac pacemakers	F1905	Device Revision or Replacement	A1304	Telemetry Discrepancy		
									A26	Insufficient Device Problem Information		
2333	2013		47270	Cardiac resynchronization therapy implantable defibrillator	CT2866	Implantable cardiac electrical stimulators			A26	Insufficient Device Problem Information		
2334	2013		47270	Cardiac resynchronization therapy implantable defibrillator	CT2866	Implantable cardiac electrical stimulators			A26	Insufficient Device Problem Information		
2335	2013		47270	Cardiac resynchronization therapy implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	F15	Recognized Device or Procedural Complication	A26	Insufficient Device Problem Information		
2336	2013		47270	Cardiac resynchronization therapy implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	F15	Recognized Device or Procedural Complication	A26	Insufficient Device Problem Information		
2337	2013		47270	Cardiac resynchronization therapy implantable defibrillator	CT2866	Implantable cardiac electrical stimulators			A26	Insufficient Device Problem Information		
2338	2013		47270	Cardiac resynchronization therapy implantable defibrillator	CT2866	Implantable cardiac electrical stimulators			A26	Insufficient Device Problem Information		
2339	2013		47270	Cardiac resynchronization therapy implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	F15	Recognized Device or Procedural Complication	A26	Insufficient Device Problem Information		
2340	2013		47270	Cardiac resynchronization therapy implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	F15	Recognized Device or Procedural Complication	A26	Insufficient Device Problem Information		
2341	2013		47270	Cardiac resynchronization therapy implantable defibrillator	CT2866	Implantable cardiac electrical stimulators			A26	Insufficient Device Problem Information		
2342	2013		47270	Cardiac resynchronization therapy implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	F21	Unexpected Deterioration	A26	Insufficient Device Problem Information		
2343	2013		47270	Cardiac resynchronization therapy implantable defibrillator	CT2866	Implantable cardiac electrical stimulators			A24	No Apparent Device Problem		
									A26	Insufficient Device Problem Information		
2344	2013		47270	Cardiac resynchronization therapy implantable defibrillator	CT2866	Implantable cardiac electrical stimulators			A26	Insufficient Device Problem Information		
2345	2013		47270	Cardiac resynchronization therapy implantable defibrillator	CT2866	Implantable cardiac electrical stimulators			A26	Insufficient Device Problem Information		
2346	2013		47270	Cardiac resynchronization therapy implantable defibrillator	CT2866	Implantable cardiac electrical stimulators			A26	Insufficient Device Problem Information		
2347	2013		35223	Endocardial/interventricular septal pacing lead	CT1684	Cardiac pacemakers and associated devices	F21	Unexpected Deterioration	A09	Output Problem		
									A090404	Output Above Specifications		
2348	2013		47270	Cardiac resynchronization therapy implantable defibrillator	CT2866	Implantable cardiac electrical stimulators			A26	Insufficient Device Problem Information		
2349	2013		47270	Cardiac resynchronization therapy implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E0202	Emotional Changes	A26	Insufficient Device Problem Information		
2350	2014	All other reportable incidents			CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A1501	Activation Problem		
									A23	Use of Device Problem		
2351	2013		47270	Cardiac resynchronization therapy implantable defibrillator	CT2866	Implantable cardiac electrical stimulators			A26	Insufficient Device Problem Information		
2352	2013				CT257	Implantable cardiac pacemakers	E0402	Hypersensitivity/Allergic Reaction	A07	Electrical /Electronic Property Problem		
									A0709	Device Sensing Problem		
2353	2013				CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A07	Electrical /Electronic Property Problem		
									A0709	Device Sensing Problem		
2354	2014				CT257	Implantable cardiac pacemakers	E2401	Insufficient Information	A07	Electrical /Electronic Property Problem	28/03/2014	
									A0708	Power Problem	28/03/2014	
2355	2013	All other reportable incidents			CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A07	Electrical /Electronic Property Problem		
									A0708	Power Problem		
2356	2013				CT257	Implantable cardiac pacemakers	F15	Recognized Device or Procedural Complication	A26	Insufficient Device Problem Information		
2357	2013		47270	Cardiac resynchronization therapy implantable defibrillator	CT2866	Implantable cardiac electrical stimulators			A26	Insufficient Device Problem Information		

2358	2013		47270	Cardiac resynchronization therapy implantable defibrillator	CT2866	Implantable cardiac electrical stimulators			A26	Insufficient Device Problem Information		
2359	2013		35223	Endocardial/interventricular septal pacing lead	CT1684	Cardiac pacemakers and associated devices	E2401	Insufficient Information	A04	Material Integrity Problem		
									A0404	Crack		
									A07	Electrical /Electronic Property Problem		
									A0709	Device Sensing Problem		
2360	2014		47270	Cardiac resynchronization therapy implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E0402	Hypersensitivity/Allergic Reaction	A26	Insufficient Device Problem Information		
2361	2013	All other reportable incidents			CT257	Implantable cardiac pacemakers	F1905	Device Revision or Replacement	A1304	Telemetry Discrepancy	24/10/2013	
									A26	Insufficient Device Problem Information	24/10/2013	
2362	2013	All other reportable incidents			CT257	Implantable cardiac pacemakers			A09	Output Problem		
									A0907	No Device Output		
									A1304	Telemetry Discrepancy		
									A26	Insufficient Device Problem Information		
2363	2014				CT257	Implantable cardiac pacemakers	F1905	Device Revision or Replacement	A07	Electrical /Electronic Property Problem	20/01/2014	
									A0708	Power Problem	20/01/2014	
2364	2013	All other reportable incidents			CT257	Implantable cardiac pacemakers	F1905	Device Revision or Replacement	A07	Electrical /Electronic Property Problem		
									A0708	Power Problem		
2365	2013		47270	Cardiac resynchronization therapy implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A07	Electrical /Electronic Property Problem		
									A0709	Device Sensing Problem		
2366	2014	All other reportable incidents			CT257	Implantable cardiac pacemakers	F1905	Device Revision or Replacement	A07	Electrical /Electronic Property Problem		
									A0708	Power Problem		
2367	2014	All other reportable incidents			CT257	Implantable cardiac pacemakers	F1905	Device Revision or Replacement	A07	Electrical /Electronic Property Problem		
									A0708	Power Problem		
2368	2014		47270	Cardiac resynchronization therapy implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A26	Insufficient Device Problem Information		
2369	2014		47270	Cardiac resynchronization therapy implantable defibrillator	CT2866	Implantable cardiac electrical stimulators			A26	Insufficient Device Problem Information		
2370	2013		47270	Cardiac resynchronization therapy implantable defibrillator	CT2866	Implantable cardiac electrical stimulators			A26	Insufficient Device Problem Information		
2371	2013		47270	Cardiac resynchronization therapy implantable defibrillator	CT2866	Implantable cardiac electrical stimulators			A26	Insufficient Device Problem Information		
2372	2013		47270	Cardiac resynchronization therapy implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E2401	Insufficient Information	A26	Insufficient Device Problem Information		
2373	2013		47263	Cardiac resynchronization therapy implantable pacemaker	CT1684	Cardiac pacemakers and associated devices	F21	Unexpected Deterioration	A26	Insufficient Device Problem Information		
					CT1685	Cardiac pacemakers	F21	Unexpected Deterioration	A26	Insufficient Device Problem Information		
					CT257	Implantable cardiac pacemakers	F21	Unexpected Deterioration	A26	Insufficient Device Problem Information		
					CT2866	Implantable cardiac electrical stimulators	F21	Unexpected Deterioration	A26	Insufficient Device Problem Information		
2374	2013				CT257	Implantable cardiac pacemakers			A07	Electrical /Electronic Property Problem		
									A0708	Power Problem		
2375	2013		47270	Cardiac resynchronization therapy implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E2401	Insufficient Information	A15	Activation, Positioning or Separation Problem		
									A150301	Separation Failure		
2376	2014				CT257	Implantable cardiac pacemakers	F1905	Device Revision or Replacement	A07	Electrical /Electronic Property Problem		
									A0708	Power Problem		
2377	2013		47270	Cardiac resynchronization therapy implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	F15	Recognized Device or Procedural Complication	A26	Insufficient Device Problem Information		
2378	2013		47270	Cardiac resynchronization therapy implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	F15	Recognized Device or Procedural Complication	A26	Insufficient Device Problem Information		
2379	2013		47270	Cardiac resynchronization therapy implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	F15	Recognized Device or Procedural Complication	A26	Insufficient Device Problem Information		
2380	2013		47270	Cardiac resynchronization therapy implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	F15	Recognized Device or Procedural Complication	A26	Insufficient Device Problem Information		
2381	2013		47270	Cardiac resynchronization therapy implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	F15	Recognized Device or Procedural Complication	A26	Insufficient Device Problem Information		
2382	2014	All other reportable incidents			CT257	Implantable cardiac pacemakers	F1905	Device Revision or Replacement	A07	Electrical /Electronic Property Problem	20/01/2014	
									A0708	Power Problem	20/01/2014	
2383	2013		47270	Cardiac resynchronization therapy implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E0506	Hemorrhage/Blood Loss/Bleeding	A24	No Apparent Device Problem		
									A26	Insufficient Device Problem Information		
2384	2013		47270	Cardiac resynchronization therapy implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	F15	Recognized Device or Procedural Complication	A26	Insufficient Device Problem Information		
2385	2013		47270	Cardiac resynchronization therapy implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E2401	Insufficient Information	A26	Insufficient Device Problem Information		
2386	2013		47270	Cardiac resynchronization therapy implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E2401	Insufficient Information	A26	Insufficient Device Problem Information		
2387	2013		47270	Cardiac resynchronization therapy implantable defibrillator	CT2866	Implantable cardiac electrical stimulators			A26	Insufficient Device Problem Information		
2388	2014				CT257	Implantable cardiac pacemakers	E0402	Hypersensitivity/Allergic Reaction	A26	Insufficient Device Problem Information		
2389	2013		47270	Cardiac resynchronization therapy implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	F14	Prolonged Episode of Care	A26	Insufficient Device Problem Information		
2390	2013		47270	Cardiac resynchronization therapy implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	F21	Unexpected Deterioration	A26	Insufficient Device Problem Information		
2391	2013		47270	Cardiac resynchronization therapy implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	F21	Unexpected Deterioration	A26	Insufficient Device Problem Information		
2392	2013		47270	Cardiac resynchronization therapy implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	F21	Unexpected Deterioration	A26	Insufficient Device Problem Information		
2393	2013		47270	Cardiac resynchronization therapy implantable defibrillator	CT2866	Implantable cardiac electrical stimulators			A26	Insufficient Device Problem Information		
2394	2013		47270	Cardiac resynchronization therapy implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E2401	Insufficient Information	A26	Insufficient Device Problem Information		
2395	2013		47270	Cardiac resynchronization therapy implantable defibrillator	CT2866	Implantable cardiac electrical stimulators			A26	Insufficient Device Problem Information		
2396	2013		47270	Cardiac resynchronization therapy implantable defibrillator	CT2866	Implantable cardiac electrical stimulators			A26	Insufficient Device Problem Information		
2397	2013		47270	Cardiac resynchronization therapy implantable defibrillator	CT2866	Implantable cardiac electrical stimulators			A26	Insufficient Device Problem Information		
2398	2013		47270	Cardiac resynchronization therapy implantable defibrillator	CT2866	Implantable cardiac electrical stimulators			A26	Insufficient Device Problem Information		
2399	2013		47270	Cardiac resynchronization therapy implantable defibrillator	CT2866	Implantable cardiac electrical stimulators			A26	Insufficient Device Problem Information		
2400	2013		47270	Cardiac resynchronization therapy implantable defibrillator	CT2866	Implantable cardiac electrical stimulators			A26	Insufficient Device Problem Information		
2401	2013		47270	Cardiac resynchronization therapy implantable defibrillator	CT2866	Implantable cardiac electrical stimulators			A26	Insufficient Device Problem Information		
2402	2014				CT257	Implantable cardiac pacemakers	E0402	Hypersensitivity/Allergic Reaction	A26	Insufficient Device Problem Information		
2403	2014		47270	Cardiac resynchronization therapy implantable	CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A07	Electrical /Electronic Property Problem		

				defibrillator					A0708	Power Problem		
2404	2014		47270	Cardiac resynchronization therapy implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E0402	Hypersensitivity/Allergic Reaction	A26	Insufficient Device Problem Information		
2405	2013	All other reportable incidents			CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A07	Electrical /Electronic Property Problem		
									A0708	Power Problem		
2406	2013		47270	Cardiac resynchronization therapy implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	F15	Recognized Device or Procedural Complication	A26	Insufficient Device Problem Information		
2407	2013		47270	Cardiac resynchronization therapy implantable defibrillator	CT2866	Implantable cardiac electrical stimulators			A26	Insufficient Device Problem Information		
2408	2013		47270	Cardiac resynchronization therapy implantable defibrillator	CT2866	Implantable cardiac electrical stimulators			A26	Insufficient Device Problem Information		
2409	2013				CT257	Implantable cardiac pacemakers	F1905	Device Revision or Replacement	A07	Electrical /Electronic Property Problem		
									A0708	Power Problem		
									A0709	Device Sensing Problem		
2410	2014				CT257	Implantable cardiac pacemakers	F1905	Device Revision or Replacement	A07	Electrical /Electronic Property Problem		
									A0708	Power Problem		
2411	2013				CT257	Implantable cardiac pacemakers	E0139	Unspecified Nervous System Problem	A09	Output Problem		
									A090404	Output Above Specifications		
									A13	Communication or Transmission Problem		
									A26	Insufficient Device Problem Information		
2412	2013		47270	Cardiac resynchronization therapy implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	F15	Recognized Device or Procedural Complication	A26	Insufficient Device Problem Information		
2413	2013		47270	Cardiac resynchronization therapy implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	F15	Recognized Device or Procedural Complication	A26	Insufficient Device Problem Information		
2414	2013		47270	Cardiac resynchronization therapy implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	F15	Recognized Device or Procedural Complication	A26	Insufficient Device Problem Information		
2415	2013				CT257	Implantable cardiac pacemakers	F1905	Device Revision or Replacement	A07	Electrical /Electronic Property Problem	02/04/2013	
									A0708	Power Problem	02/04/2013	
2416	2013				CT257	Implantable cardiac pacemakers	E2401	Insufficient Information	A07	Electrical /Electronic Property Problem		
									A0708	Power Problem		
2417	2014	All other reportable incidents			CT257	Implantable cardiac pacemakers	F1905	Device Revision or Replacement	A16	Protective Measures Problem		
									A1602	Fail-Safe Problem		
2418	2013	All other reportable incidents			CT257	Implantable cardiac pacemakers	F1905	Device Revision or Replacement	A07	Electrical /Electronic Property Problem		
									A0708	Power Problem		
2419	2013	All other reportable incidents			CT257	Implantable cardiac pacemakers	F1905	Device Revision or Replacement	A07	Electrical /Electronic Property Problem	01/11/2013	
									A0708	Power Problem	01/11/2013	
2420	2013	All other reportable incidents	35223	Endocardial/interventricular septal pacing lead	CT1684	Cardiac pacemakers and associated devices	F1905	Device Revision or Replacement	A07	Electrical /Electronic Property Problem		
									A0709	Device Sensing Problem		
2421	2014		47270	Cardiac resynchronization therapy implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E0402	Hypersensitivity/Allergic Reaction	A26	Insufficient Device Problem Information		
2422	2014	All other reportable incidents	35223	Endocardial/interventricular septal pacing lead	CT1684	Cardiac pacemakers and associated devices	F1905	Device Revision or Replacement	A07	Electrical /Electronic Property Problem	17/07/2014	
									A0709	Device Sensing Problem	17/07/2014	
2423	2014		47270	Cardiac resynchronization therapy implantable defibrillator	CT2866	Implantable cardiac electrical stimulators			A26	Insufficient Device Problem Information		
2424	2014	All other reportable incidents			CT257	Implantable cardiac pacemakers	F1905	Device Revision or Replacement	A07	Electrical /Electronic Property Problem		
									A0708	Power Problem		
									A1304	Telemetry Discrepancy		
									A26	Insufficient Device Problem Information		
2425	2013		47270	Cardiac resynchronization therapy implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	F14	Prolonged Episode of Care	A09	Output Problem		
									A090404	Output Above Specifications		
2426	2013				CT257	Implantable cardiac pacemakers	F1905	Device Revision or Replacement	A07	Electrical /Electronic Property Problem		
									A0708	Power Problem		
									A0709	Device Sensing Problem		
2427	2013		47270	Cardiac resynchronization therapy implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E0402	Hypersensitivity/Allergic Reaction	A26	Insufficient Device Problem Information		
2428	2013	All other reportable incidents			CT257	Implantable cardiac pacemakers	F1905	Device Revision or Replacement	A07	Electrical /Electronic Property Problem	24/12/2012	
									A0708	Power Problem	24/12/2012	
2429	2014				CT257	Implantable cardiac pacemakers	E0402	Hypersensitivity/Allergic Reaction	A26	Insufficient Device Problem Information		
2430	2013		47270	Cardiac resynchronization therapy implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A07	Electrical /Electronic Property Problem		
									A0709	Device Sensing Problem		
2431	2013				CT257	Implantable cardiac pacemakers	F15	Recognized Device or Procedural Complication	A26	Insufficient Device Problem Information		
2432	2013				CT257	Implantable cardiac pacemakers	F15	Recognized Device or Procedural Complication	A26	Insufficient Device Problem Information		
2433	2013		47270	Cardiac resynchronization therapy implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	F15	Recognized Device or Procedural Complication	A26	Insufficient Device Problem Information		
2434	2014	All other reportable incidents			CT257	Implantable cardiac pacemakers	F1905	Device Revision or Replacement	A07	Electrical /Electronic Property Problem	24/03/2014	
									A0708	Power Problem	24/03/2014	
2435	2013		47270	Cardiac resynchronization therapy implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E0402	Hypersensitivity/Allergic Reaction	A26	Insufficient Device Problem Information		
2436	2013		47270	Cardiac resynchronization therapy implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	F14	Prolonged Episode of Care	A26	Insufficient Device Problem Information		
2437	2013		47270	Cardiac resynchronization therapy implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	F21	Unexpected Deterioration	A26	Insufficient Device Problem Information		
2438	2013		47270	Cardiac resynchronization therapy implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	F15	Recognized Device or Procedural Complication	A26	Insufficient Device Problem Information		
2439	2013		47270	Cardiac resynchronization therapy implantable defibrillator	CT2866	Implantable cardiac electrical stimulators			A26	Insufficient Device Problem Information		
2440	2013		47270	Cardiac resynchronization therapy implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	F21	Unexpected Deterioration	A26	Insufficient Device Problem Information		
2441	2013		47270	Cardiac resynchronization therapy implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	F21	Unexpected Deterioration	A26	Insufficient Device Problem Information		
2442	2013		47270	Cardiac resynchronization therapy implantable defibrillator	CT2866	Implantable cardiac electrical stimulators			A26	Insufficient Device Problem Information		
2443	2013		47270	Cardiac resynchronization therapy implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	F15	Recognized Device or Procedural Complication	A26	Insufficient Device Problem Information		
2444	2013		47270	Cardiac resynchronization therapy implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	F15	Recognized Device or Procedural Complication	A26	Insufficient Device Problem Information		
2445	2013		47270	Cardiac resynchronization therapy implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	F14	Prolonged Episode of Care	A26	Insufficient Device Problem Information		
2446	2013		47270	Cardiac resynchronization therapy implantable defibrillator	CT2866	Implantable cardiac electrical stimulators			A26	Insufficient Device Problem Information		
2447	2013		47270	Cardiac resynchronization therapy implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	F21	Unexpected Deterioration	A26	Insufficient Device Problem Information		
2448	2013		47270	Cardiac resynchronization therapy implantable defibrillator	CT2866	Implantable cardiac electrical stimulators			A26	Insufficient Device Problem Information		

2449	2013		47270	Cardiac resynchronization therapy implantable defibrillator	CT2866	Implantable cardiac electrical stimulators			A26	Insufficient Device Problem Information		
2450	2013		47270	Cardiac resynchronization therapy implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	F14	Prolonged Episode of Care	A26	Insufficient Device Problem Information		
2451	2013		47270	Cardiac resynchronization therapy implantable defibrillator	CT2866	Implantable cardiac electrical stimulators			A26	Insufficient Device Problem Information		
2452	2013		47270	Cardiac resynchronization therapy implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	F21	Unexpected Deterioration	A26	Insufficient Device Problem Information		
2453	2013		47270	Cardiac resynchronization therapy implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	F15	Recognized Device or Procedural Complication	A26	Insufficient Device Problem Information		
2454	2014		47270	Cardiac resynchronization therapy implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	F15	Recognized Device or Procedural Complication	A26	Insufficient Device Problem Information		
2455	2014		47270	Cardiac resynchronization therapy implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	F15	Recognized Device or Procedural Complication	A26	Insufficient Device Problem Information		
2456	2013		47270	Cardiac resynchronization therapy implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	F14	Prolonged Episode of Care	A26	Insufficient Device Problem Information		
2457	2013		47270	Cardiac resynchronization therapy implantable defibrillator	CT2866	Implantable cardiac electrical stimulators			A26	Insufficient Device Problem Information		
2458	2013		47270	Cardiac resynchronization therapy implantable defibrillator	CT2866	Implantable cardiac electrical stimulators			A26	Insufficient Device Problem Information		
2459	2013		47270	Cardiac resynchronization therapy implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E2401	Insufficient Information	A26	Insufficient Device Problem Information		
2460	2013		47270	Cardiac resynchronization therapy implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E0402	Hypersensitivity/Allergic Reaction	A26	Insufficient Device Problem Information		
2461	2014				CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A16	Protective Measures Problem		
									A1602	Fail-Safe Problem		
2462	2014		47270	Cardiac resynchronization therapy implantable defibrillator	CT2866	Implantable cardiac electrical stimulators			A26	Insufficient Device Problem Information		
2463	2014				CT257	Implantable cardiac pacemakers	F1905	Device Revision or Replacement	A13	Communication or Transmission Problem		
									A26	Insufficient Device Problem Information		
2464	2014		47270	Cardiac resynchronization therapy implantable defibrillator	CT2866	Implantable cardiac electrical stimulators			A26	Insufficient Device Problem Information		
2465	2013		47270	Cardiac resynchronization therapy implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	F1905	Device Revision or Replacement	A05	Mechanical Problem	05/06/2013	
									A0506	Mechanical Jam	05/06/2013	
2466	2014		47270	Cardiac resynchronization therapy implantable defibrillator	CT2866	Implantable cardiac electrical stimulators			A26	Insufficient Device Problem Information		
2467	2013		47270	Cardiac resynchronization therapy implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A07	Electrical /Electronic Property Problem		
									A0709	Device Sensing Problem		
									A09	Output Problem		
									A090404	Output Above Specifications		
2468	2013	All other reportable incidents	47270	Cardiac resynchronization therapy implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	F1905	Device Revision or Replacement	A07	Electrical /Electronic Property Problem	22/10/2013	
									A0708	Power Problem	22/10/2013	
2469	2013		47270	Cardiac resynchronization therapy implantable defibrillator	CT2866	Implantable cardiac electrical stimulators			A24	No Apparent Device Problem		
									A26	Insufficient Device Problem Information		
2470	2013				CT257	Implantable cardiac pacemakers	F1905	Device Revision or Replacement	A07	Electrical /Electronic Property Problem		
									A0709	Device Sensing Problem		
2471	2013		47270	Cardiac resynchronization therapy implantable defibrillator	CT2866	Implantable cardiac electrical stimulators			A26	Insufficient Device Problem Information		
2472	2013		47270	Cardiac resynchronization therapy implantable defibrillator	CT2866	Implantable cardiac electrical stimulators			A26	Insufficient Device Problem Information		
2473	2013		47270	Cardiac resynchronization therapy implantable defibrillator	CT2866	Implantable cardiac electrical stimulators			A26	Insufficient Device Problem Information		
2474	2014	All other reportable incidents			CT257	Implantable cardiac pacemakers	F1905	Device Revision or Replacement	A07	Electrical /Electronic Property Problem	28/07/2014	
									A0708	Power Problem	28/07/2014	
2475	2013		47270	Cardiac resynchronization therapy implantable defibrillator	CT2866	Implantable cardiac electrical stimulators			A26	Insufficient Device Problem Information		
2476	2014		47270	Cardiac resynchronization therapy implantable defibrillator	CT2866	Implantable cardiac electrical stimulators			A07	Electrical /Electronic Property Problem		
									A0708	Power Problem		
2477	2013	All other reportable incidents			CT257	Implantable cardiac pacemakers	F1905	Device Revision or Replacement	A09	Output Problem		
									A090404	Output Above Specifications		
2478	2013	All other reportable incidents			CT257	Implantable cardiac pacemakers			A07	Electrical /Electronic Property Problem		
									A0708	Power Problem		
2479	2013	All other reportable incidents			CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A1501	Activation Problem		
									A23	Use of Device Problem		
2480	2014				CT257	Implantable cardiac pacemakers	F1905	Device Revision or Replacement	A07	Electrical /Electronic Property Problem		
									A0708	Power Problem		
2481	2013		47270	Cardiac resynchronization therapy implantable defibrillator	CT2866	Implantable cardiac electrical stimulators			A26	Insufficient Device Problem Information		
2482	2013		47270	Cardiac resynchronization therapy implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A07	Electrical /Electronic Property Problem		
									A0709	Device Sensing Problem		
2483	2014				CT257	Implantable cardiac pacemakers	E2401	Insufficient Information	A1304	Telemetry Discrepancy		
									A26	Insufficient Device Problem Information		
2484	2013		47270	Cardiac resynchronization therapy implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	F15	Recognized Device or Procedural Complication	A26	Insufficient Device Problem Information		
2485	2014		47270	Cardiac resynchronization therapy implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A07	Electrical /Electronic Property Problem		
									A0708	Power Problem		
2486	2014				CT257	Implantable cardiac pacemakers	F21	Unexpected Deterioration	A26	Insufficient Device Problem Information		
2487	2014				CT257	Implantable cardiac pacemakers	F1905	Device Revision or Replacement	A1304	Telemetry Discrepancy		
									A26	Insufficient Device Problem Information		
2488	2014		47270	Cardiac resynchronization therapy implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	F14	Prolonged Episode of Care	A07	Electrical /Electronic Property Problem		
									A0709	Device Sensing Problem		
2489	2014		47270	Cardiac resynchronization therapy implantable defibrillator	CT2866	Implantable cardiac electrical stimulators			A07	Electrical /Electronic Property Problem		
									A0708	Power Problem		
2490	2014		47270	Cardiac resynchronization therapy implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E0402	Hypersensitivity/Allergic Reaction	A26	Insufficient Device Problem Information		
2491	2014		47270	Cardiac resynchronization therapy implantable defibrillator	CT2866	Implantable cardiac electrical stimulators			A26	Insufficient Device Problem Information		
2492	2013	All other reportable incidents			CT257	Implantable cardiac pacemakers	F14	Prolonged Episode of Care	A09	Output Problem		
									A0907	No Device Output		
									A1304	Telemetry Discrepancy		
									A26	Insufficient Device Problem Information		
2493	2013		47270	Cardiac resynchronization therapy implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E2401	Insufficient Information	A26	Insufficient Device Problem Information		

2494	2013		47270	Cardiac resynchronization therapy implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E2401	Insufficient Information	A26	Insufficient Device Problem Information		
2495	2014				CT257	Implantable cardiac pacemakers	F21	Unexpected Deterioration	A07	Electrical /Electronic Property Problem		
									A0709	Device Sensing Problem		
									A16	Protective Measures Problem		
									A1602	Fail-Safe Problem		
2496	2013		47270	Cardiac resynchronization therapy implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A1304	Telemetry Discrepancy		
									A26	Insufficient Device Problem Information		
2497	2013		47270	Cardiac resynchronization therapy implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A1304	Telemetry Discrepancy		
									A26	Insufficient Device Problem Information		
2498	2014		35223	Endocardial/interventricular septal pacing lead	CT1684	Cardiac pacemakers and associated devices	F14	Prolonged Episode of Care	A15	Activation, Positioning or Separation Problem		
									A150203	Difficult or Delayed Positioning		
2499	2014		47270	Cardiac resynchronization therapy implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E0402	Hypersensitivity/Allergic Reaction	A26	Insufficient Device Problem Information		
2500	2014	All other reportable incidents			CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A1501	Activation Problem		
									A23	Use of Device Problem		
2501	2013	All other reportable incidents			CT257	Implantable cardiac pacemakers	E2401	Insufficient Information	A1304	Telemetry Discrepancy	15/02/2013	
									A26	Insufficient Device Problem Information	15/02/2013	
2502	2014		47270	Cardiac resynchronization therapy implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A07	Electrical /Electronic Property Problem		
									A0708	Power Problem		
2503	2014				CT257	Implantable cardiac pacemakers	F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A09	Output Problem		
									A0907	No Device Output		
2504	2014	All other reportable incidents			CT257	Implantable cardiac pacemakers	F1905	Device Revision or Replacement	A07	Electrical /Electronic Property Problem		
									A0709	Device Sensing Problem		
2505	2014				CT257	Implantable cardiac pacemakers	F1905	Device Revision or Replacement	A05	Mechanical Problem		
									A0512	Unintended Movement		
2506	2013				CT257	Implantable cardiac pacemakers	F14	Prolonged Episode of Care	A26	Insufficient Device Problem Information		
2507	2013	All other reportable incidents	35223	Endocardial/interventricular septal pacing lead	CT1684	Cardiac pacemakers and associated devices	E1906	Unspecified Infection	A04	Material Integrity Problem	30/05/2013	
									A0404	Crack	30/05/2013	
2508	2014				CT257	Implantable cardiac pacemakers	F21	Unexpected Deterioration	A26	Insufficient Device Problem Information		
2509	2013				CT257	Implantable cardiac pacemakers	F14	Prolonged Episode of Care	A26	Insufficient Device Problem Information		
2510	2013				CT257	Implantable cardiac pacemakers	F14	Prolonged Episode of Care	A26	Insufficient Device Problem Information		
2511	2014		47270	Cardiac resynchronization therapy implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E2104	Electric Shock	A07	Electrical /Electronic Property Problem		
									A0709	Device Sensing Problem		
2512	2013				CT257	Implantable cardiac pacemakers	F14	Prolonged Episode of Care	A26	Insufficient Device Problem Information		
2513	2013		47270	Cardiac resynchronization therapy implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E0506	Hemorrhage/Blood Loss/Bleeding	A24	No Apparent Device Problem		
									A26	Insufficient Device Problem Information		
2514	2014				CT257	Implantable cardiac pacemakers	F1905	Device Revision or Replacement	A07	Electrical /Electronic Property Problem	11/04/2014	
									A0708	Power Problem	11/04/2014	
									A09	Output Problem	11/04/2014	
									A0907	No Device Output	11/04/2014	
2515	2013		47270	Cardiac resynchronization therapy implantable defibrillator	CT2866	Implantable cardiac electrical stimulators			A24	No Apparent Device Problem		
									A26	Insufficient Device Problem Information		
2516	2013		47270	Cardiac resynchronization therapy implantable defibrillator	CT2866	Implantable cardiac electrical stimulators			A24	No Apparent Device Problem		
									A26	Insufficient Device Problem Information		
2517	2013		47270	Cardiac resynchronization therapy implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	F21	Unexpected Deterioration	A26	Insufficient Device Problem Information		
2518	2013		47270	Cardiac resynchronization therapy implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	F15	Recognized Device or Procedural Complication	A26	Insufficient Device Problem Information		
2519	2013		47270	Cardiac resynchronization therapy implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	F21	Unexpected Deterioration	A26	Insufficient Device Problem Information		
2520	2014	All other reportable incidents			CT257	Implantable cardiac pacemakers	F1905	Device Revision or Replacement	A07	Electrical /Electronic Property Problem		
									A0708	Power Problem		
2521	2014		35223	Endocardial/interventricular septal pacing lead	CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A15	Activation, Positioning or Separation Problem		
									A150203	Difficult or Delayed Positioning		
2522	2014	All other reportable incidents			CT257	Implantable cardiac pacemakers	F1905	Device Revision or Replacement	A07	Electrical /Electronic Property Problem	19/03/2014	
									A0708	Power Problem	19/03/2014	
2523	2013		47270	Cardiac resynchronization therapy implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E0402	Hypersensitivity/Allergic Reaction	A26	Insufficient Device Problem Information		
2524	2014	Death			CT257	Implantable cardiac pacemakers	F21	Unexpected Deterioration	A24	No Apparent Device Problem		
									A26	Insufficient Device Problem Information		
2525	2013		47270	Cardiac resynchronization therapy implantable defibrillator	CT2866	Implantable cardiac electrical stimulators			A26	Insufficient Device Problem Information		
2526	2013		47270	Cardiac resynchronization therapy implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E2401	Insufficient Information	A26	Insufficient Device Problem Information		
2527	2013		47270	Cardiac resynchronization therapy implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	F15	Recognized Device or Procedural Complication	A26	Insufficient Device Problem Information		
2528	2013		47270	Cardiac resynchronization therapy implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A26	Insufficient Device Problem Information		
2529	2013		47270	Cardiac resynchronization therapy implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A26	Insufficient Device Problem Information		
2530	2013		47270	Cardiac resynchronization therapy implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A26	Insufficient Device Problem Information		
2531	2013		47270	Cardiac resynchronization therapy implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A26	Insufficient Device Problem Information		
2532	2013		47270	Cardiac resynchronization therapy implantable defibrillator	CT2866	Implantable cardiac electrical stimulators			A26	Insufficient Device Problem Information		
2533	2013		47270	Cardiac resynchronization therapy implantable defibrillator	CT2866	Implantable cardiac electrical stimulators			A26	Insufficient Device Problem Information		
2534	2013		47270	Cardiac resynchronization therapy implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	F15	Recognized Device or Procedural Complication	A26	Insufficient Device Problem Information		
2535	2013		47270	Cardiac resynchronization therapy implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	F15	Recognized Device or Procedural Complication	A26	Insufficient Device Problem Information		
2536	2013		47270	Cardiac resynchronization therapy implantable defibrillator	CT2866	Implantable cardiac electrical stimulators			A26	Insufficient Device Problem Information		
2537	2013		47270	Cardiac resynchronization therapy implantable defibrillator	CT2866	Implantable cardiac electrical stimulators			A26	Insufficient Device Problem Information		
2538	2013		47270	Cardiac resynchronization therapy implantable defibrillator	CT2866	Implantable cardiac electrical stimulators			A26	Insufficient Device Problem Information		
2539	2013		47270	Cardiac resynchronization therapy implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	F15	Recognized Device or Procedural Complication	A26	Insufficient Device Problem Information		
2540	2014		47270	Cardiac resynchronization therapy implantable defibrillator	CT2866	Implantable cardiac electrical stimulators			A26	Insufficient Device Problem Information		

2541	2014		47270	Cardiac resynchronization therapy implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E0402	Hypersensitivity/Allergic Reaction	A26	Insufficient Device Problem Information		
2542	2013				CT257	Implantable cardiac pacemakers	F1905	Device Revision or Replacement	A12	Connection Problem		
									A1205	Loose or Intermittent Connection		
2543	2014	Death			CT257	Implantable cardiac pacemakers	F21	Unexpected Deterioration	A26	Insufficient Device Problem Information		
2544	2014	All other reportable incidents			CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A1104	Computer Operating System Problem		
2545	2014		47270	Cardiac resynchronization therapy implantable defibrillator	CT2866	Implantable cardiac electrical stimulators			A26	Insufficient Device Problem Information		
2546	2013		47270	Cardiac resynchronization therapy implantable defibrillator	CT2866	Implantable cardiac electrical stimulators			A26	Insufficient Device Problem Information		
2547	2014		47270	Cardiac resynchronization therapy implantable defibrillator	CT2866	Implantable cardiac electrical stimulators			A26	Insufficient Device Problem Information		
2548	2014		47270	Cardiac resynchronization therapy implantable defibrillator	CT2866	Implantable cardiac electrical stimulators			A26	Insufficient Device Problem Information		
2549	2014				CT257	Implantable cardiac pacemakers	F1905	Device Revision or Replacement	A07	Electrical /Electronic Property Problem		
									A0708	Power Problem		
2550	2014		47270	Cardiac resynchronization therapy implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	F14	Prolonged Episode of Care	A07	Electrical /Electronic Property Problem		
									A0708	Power Problem		
2551	2014		47270	Cardiac resynchronization therapy implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	F21	Unexpected Deterioration	A26	Insufficient Device Problem Information		
2552	2014				CT257	Implantable cardiac pacemakers			A26	Insufficient Device Problem Information		
2553	2013		47270	Cardiac resynchronization therapy implantable defibrillator	CT2866	Implantable cardiac electrical stimulators			A26	Insufficient Device Problem Information		
2554	2014	All other reportable incidents			CT257	Implantable cardiac pacemakers	F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A07	Electrical /Electronic Property Problem	21/08/2014	
									A0708	Power Problem	21/08/2014	
2555	2013		47270	Cardiac resynchronization therapy implantable defibrillator	CT2866	Implantable cardiac electrical stimulators			A26	Insufficient Device Problem Information		
2556	2013		47270	Cardiac resynchronization therapy implantable defibrillator	CT2866	Implantable cardiac electrical stimulators			A26	Insufficient Device Problem Information		
2557	2014				CT257	Implantable cardiac pacemakers	F21	Unexpected Deterioration	A26	Insufficient Device Problem Information		
2558	2014		47270	Cardiac resynchronization therapy implantable defibrillator	CT2866	Implantable cardiac electrical stimulators			A26	Insufficient Device Problem Information		
2559	2014				CT257	Implantable cardiac pacemakers			A07	Electrical /Electronic Property Problem		
									A0709	Device Sensing Problem		
2560	2014				CT257	Implantable cardiac pacemakers	E0402	Hypersensitivity/Allergic Reaction	A26	Insufficient Device Problem Information		
2561	2013		35223	Endocardial/interventricular septal pacing lead	CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A07	Electrical /Electronic Property Problem		
									A0709	Device Sensing Problem		
2562	2013		47270	Cardiac resynchronization therapy implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	F1905	Device Revision or Replacement	A07	Electrical /Electronic Property Problem		
									A0708	Power Problem		
2563	2014	All other reportable incidents			CT257	Implantable cardiac pacemakers	F1905	Device Revision or Replacement	A09	Output Problem	07/01/2014	
									A0907	No Device Output	07/01/2014	
2564	2013	All other reportable incidents			CT257	Implantable cardiac pacemakers	F1905	Device Revision or Replacement	A07	Electrical /Electronic Property Problem		
									A0708	Power Problem		
2565	2013		47270	Cardiac resynchronization therapy implantable defibrillator	CT2866	Implantable cardiac electrical stimulators			A26	Insufficient Device Problem Information		
2566	2013		47270	Cardiac resynchronization therapy implantable defibrillator	CT2866	Implantable cardiac electrical stimulators			A26	Insufficient Device Problem Information		
2567	2013				CT257	Implantable cardiac pacemakers	F21	Unexpected Deterioration	A07	Electrical /Electronic Property Problem		
									A0709	Device Sensing Problem		
2568	2013		47270	Cardiac resynchronization therapy implantable defibrillator	CT2866	Implantable cardiac electrical stimulators			A26	Insufficient Device Problem Information		
2569	2013		47270	Cardiac resynchronization therapy implantable defibrillator	CT2866	Implantable cardiac electrical stimulators			A26	Insufficient Device Problem Information		
2570	2013		47270	Cardiac resynchronization therapy implantable defibrillator	CT2866	Implantable cardiac electrical stimulators			A26	Insufficient Device Problem Information		
2571	2013		47270	Cardiac resynchronization therapy implantable defibrillator	CT2866	Implantable cardiac electrical stimulators			A26	Insufficient Device Problem Information		
2572	2013				CT257	Implantable cardiac pacemakers	F1905	Device Revision or Replacement	A07	Electrical /Electronic Property Problem		
									A0708	Power Problem		
2573	2013		47270	Cardiac resynchronization therapy implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E1906	Unspecified Infection	A24	No Apparent Device Problem		
									A26	Insufficient Device Problem Information		
2574	2013		47270	Cardiac resynchronization therapy implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E2401	Insufficient Information	A24	No Apparent Device Problem		
									A26	Insufficient Device Problem Information		
2575	2013		47270	Cardiac resynchronization therapy implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	F21	Unexpected Deterioration	A24	No Apparent Device Problem		
									A26	Insufficient Device Problem Information		
2576	2013		47270	Cardiac resynchronization therapy implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	F21	Unexpected Deterioration	A24	No Apparent Device Problem		
									A26	Insufficient Device Problem Information		
2577	2013		47270	Cardiac resynchronization therapy implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	F15	Recognized Device or Procedural Complication	A26	Insufficient Device Problem Information		
2578	2013		47270	Cardiac resynchronization therapy implantable defibrillator	CT2866	Implantable cardiac electrical stimulators			A24	No Apparent Device Problem		
									A26	Insufficient Device Problem Information		
2579	2013		47270	Cardiac resynchronization therapy implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E1906	Unspecified Infection	A24	No Apparent Device Problem		
									A26	Insufficient Device Problem Information		
2580	2013		47270	Cardiac resynchronization therapy implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E0402	Hypersensitivity/Allergic Reaction	A26	Insufficient Device Problem Information		
2581	2013		47270	Cardiac resynchronization therapy implantable defibrillator	CT2866	Implantable cardiac electrical stimulators			A24	No Apparent Device Problem		
									A26	Insufficient Device Problem Information		
2582	2013		47270	Cardiac resynchronization therapy implantable defibrillator	CT2866	Implantable cardiac electrical stimulators			A24	No Apparent Device Problem		
									A26	Insufficient Device Problem Information		
2583	2013		47270	Cardiac resynchronization therapy implantable defibrillator	CT2866	Implantable cardiac electrical stimulators			A0203	Defective Device		
									A26	Insufficient Device Problem Information		
2584	2014		47270	Cardiac resynchronization therapy implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E0402	Hypersensitivity/Allergic Reaction	A26	Insufficient Device Problem Information		
2585	2013				CT257	Implantable cardiac pacemakers	F14	Prolonged Episode of Care	A07	Electrical /Electronic Property Problem		
									A0708	Power Problem		
2586	2013				CT257	Implantable cardiac pacemakers			A07	Electrical /Electronic Property Problem		
									A0708	Power Problem		
2587	2014	All other reportable incidents	35223	Endocardial/interventricular septal pacing lead	CT1684	Cardiac pacemakers and associated devices	F1905	Device Revision or Replacement	A07	Electrical /Electronic Property Problem	02/09/2014	
									A0709	Device Sensing Problem	02/09/2014	
2588	2013	All other reportable incidents			CT257	Implantable cardiac pacemakers	F1905	Device Revision or Replacement	A07	Electrical /Electronic Property Problem		
									A0708	Power Problem		
2589	2013				CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A09	Output Problem		

									A0907	No Device Output		
									A15	Activation, Positioning or Separation Problem		
									A150101	Activation Failure		
2590	2014		47270	Cardiac resynchronization therapy implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E2401	Insufficient Information	A26	Insufficient Device Problem Information		
2591	2014				CT257	Implantable cardiac pacemakers			A26	Insufficient Device Problem Information		
2592	2013		47270	Cardiac resynchronization therapy implantable defibrillator	CT2866	Implantable cardiac electrical stimulators			A26	Insufficient Device Problem Information		
2593	2013		47270	Cardiac resynchronization therapy implantable defibrillator	CT2866	Implantable cardiac electrical stimulators			A26	Insufficient Device Problem Information		
2594	2013		47270	Cardiac resynchronization therapy implantable defibrillator	CT2866	Implantable cardiac electrical stimulators			A26	Insufficient Device Problem Information		
2595	2013		47270	Cardiac resynchronization therapy implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	F15	Recognized Device or Procedural Complication	A26	Insufficient Device Problem Information		
2596	2014				CT257	Implantable cardiac pacemakers			A26	Insufficient Device Problem Information		
2597	2013		47270	Cardiac resynchronization therapy implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	F15	Recognized Device or Procedural Complication	A26	Insufficient Device Problem Information		
2598	2013		47270	Cardiac resynchronization therapy implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	F15	Recognized Device or Procedural Complication	A26	Insufficient Device Problem Information		
2599	2013		47270	Cardiac resynchronization therapy implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	F15	Recognized Device or Procedural Complication	A26	Insufficient Device Problem Information		
2600	2013				CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A07	Electrical /Electronic Property Problem		
									A0708	Power Problem		
2601	2013		47270	Cardiac resynchronization therapy implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	F21	Unexpected Deterioration	A24	No Apparent Device Problem		
									A26	Insufficient Device Problem Information		
2602	2013		47270	Cardiac resynchronization therapy implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A07	Electrical /Electronic Property Problem		
									A0709	Device Sensing Problem		
2603	2013		47270	Cardiac resynchronization therapy implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	F21	Unexpected Deterioration	A01	Patient Device Interaction Problem		
									A17	Compatibility Problem		
2604	2013		47270	Cardiac resynchronization therapy implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	F15	Recognized Device or Procedural Complication	A26	Insufficient Device Problem Information		
2605	2013		47270	Cardiac resynchronization therapy implantable defibrillator	CT2866	Implantable cardiac electrical stimulators			A26	Insufficient Device Problem Information		
2606	2013		47270	Cardiac resynchronization therapy implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E1906	Unspecified Infection	A24	No Apparent Device Problem		
									A26	Insufficient Device Problem Information		
2607	2013		47270	Cardiac resynchronization therapy implantable defibrillator	CT2866	Implantable cardiac electrical stimulators			A26	Insufficient Device Problem Information		
2608	2013		47270	Cardiac resynchronization therapy implantable defibrillator	CT2866	Implantable cardiac electrical stimulators			A26	Insufficient Device Problem Information		
2609	2013		47270	Cardiac resynchronization therapy implantable defibrillator	CT2866	Implantable cardiac electrical stimulators			A26	Insufficient Device Problem Information		
2610	2013		47270	Cardiac resynchronization therapy implantable defibrillator	CT2866	Implantable cardiac electrical stimulators			A26	Insufficient Device Problem Information		
2611	2013		47270	Cardiac resynchronization therapy implantable defibrillator	CT2866	Implantable cardiac electrical stimulators			A26	Insufficient Device Problem Information		
2612	2013		47270	Cardiac resynchronization therapy implantable defibrillator	CT2866	Implantable cardiac electrical stimulators			A26	Insufficient Device Problem Information		
2613	2013		47270	Cardiac resynchronization therapy implantable defibrillator	CT2866	Implantable cardiac electrical stimulators			A24	No Apparent Device Problem		
									A26	Insufficient Device Problem Information		
2614	2013		47270	Cardiac resynchronization therapy implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E2401	Insufficient Information	A24	No Apparent Device Problem		
									A26	Insufficient Device Problem Information		
2615	2014	All other reportable incidents			CT257	Implantable cardiac pacemakers	F1905	Device Revision or Replacement	A17	Compatibility Problem		
									A1701	Component or Accessory Incompatibility		
2616	2013				CT257	Implantable cardiac pacemakers	F15	Recognized Device or Procedural Complication	A26	Insufficient Device Problem Information		
2617	2013	All other reportable incidents			CT257	Implantable cardiac pacemakers	F1905	Device Revision or Replacement	A1304	Telemetry Discrepancy		
									A26	Insufficient Device Problem Information		
2618	2014		35852	Single-chamber implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E1906	Unspecified Infection	A01	Patient Device Interaction Problem		
									A17	Compatibility Problem		
2619	2013		47270	Cardiac resynchronization therapy implantable defibrillator	CT2866	Implantable cardiac electrical stimulators			A26	Insufficient Device Problem Information		
2620	2014		47270	Cardiac resynchronization therapy implantable defibrillator	CT2866	Implantable cardiac electrical stimulators			A26	Insufficient Device Problem Information		
2621	2014		47270	Cardiac resynchronization therapy implantable defibrillator	CT2866	Implantable cardiac electrical stimulators			A26	Insufficient Device Problem Information		
2622	2013		47270	Cardiac resynchronization therapy implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A07	Electrical /Electronic Property Problem		
									A0709	Device Sensing Problem		
2623	2013				CT257	Implantable cardiac pacemakers	E0402	Hypersensitivity/Allergic Reaction	A07	Electrical /Electronic Property Problem		
									A0708	Power Problem		
2624	2014				CT257	Implantable cardiac pacemakers	F1905	Device Revision or Replacement	A07	Electrical /Electronic Property Problem		
									A0708	Power Problem		
2625	2014	All other reportable incidents	47270	Cardiac resynchronization therapy implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	F1905	Device Revision or Replacement	A07	Electrical /Electronic Property Problem	16/09/2012	
									A0708	Power Problem	16/09/2012	
2626	2014	All other reportable incidents			CT257	Implantable cardiac pacemakers	F1905	Device Revision or Replacement	A07	Electrical /Electronic Property Problem	11/04/2014	
									A0708	Power Problem	11/04/2014	
									A09	Output Problem	11/04/2014	
									A0907	No Device Output	11/04/2014	
2627	2013		47270	Cardiac resynchronization therapy implantable defibrillator	CT2866	Implantable cardiac electrical stimulators			A07	Electrical /Electronic Property Problem		
									A0709	Device Sensing Problem		
2628	2013		47270	Cardiac resynchronization therapy implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	F1905	Device Revision or Replacement	A24	No Apparent Device Problem		
									A26	Insufficient Device Problem Information		
2629	2014		47270	Cardiac resynchronization therapy implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	F1905	Device Revision or Replacement	A07	Electrical /Electronic Property Problem		
									A0708	Power Problem		
2630	2013		47270	Cardiac resynchronization therapy implantable defibrillator	CT2866	Implantable cardiac electrical stimulators			A26	Insufficient Device Problem Information		
2631	2013		47270	Cardiac resynchronization therapy implantable defibrillator	CT2866	Implantable cardiac electrical stimulators			A26	Insufficient Device Problem Information		
2632	2013		36104	Implantable pacemaker analyser	CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A09	Output Problem		
									A0907	No Device Output		
2633	2013				CT257	Implantable cardiac pacemakers	E2401	Insufficient Information	A24	No Apparent Device Problem		
									A26	Insufficient Device Problem Information		
2634	2013		47270	Cardiac resynchronization therapy implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A07	Electrical /Electronic Property Problem		
									A0709	Device Sensing Problem		
2635	2013	All other reportable incidents			CT257	Implantable cardiac pacemakers	F1905	Device Revision or Replacement	A07	Electrical /Electronic Property Problem	08/07/2013	

									A0708	Power Problem	08/07/2013	
2636	2013		47270	Cardiac resynchronization therapy implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	F15	Recognized Device or Procedural Complication	A26	Insufficient Device Problem Information		
2637	2013				CT257	Implantable cardiac pacemakers	F15	Recognized Device or Procedural Complication	A26	Insufficient Device Problem Information		
2638	2013				CT257	Implantable cardiac pacemakers	E2401	Insufficient Information	A12	Connection Problem		
2639	2013		47270	Cardiac resynchronization therapy implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	F14	Prolonged Episode of Care	A07	Electrical /Electronic Property Problem		
									A0708	Power Problem		
2640	2014	All other reportable incidents			CT257	Implantable cardiac pacemakers	F14	Prolonged Episode of Care	A07	Electrical /Electronic Property Problem		
									A0708	Power Problem		
2641	2013	All other reportable incidents	47270	Cardiac resynchronization therapy implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	F1905	Device Revision or Replacement	A07	Electrical /Electronic Property Problem	22/12/2012	
									A0708	Power Problem	22/12/2012	
2642	2013		47270	Cardiac resynchronization therapy implantable defibrillator	CT2866	Implantable cardiac electrical stimulators			A26	Insufficient Device Problem Information		
2643	2014		47270	Cardiac resynchronization therapy implantable defibrillator	CT2866	Implantable cardiac electrical stimulators			A26	Insufficient Device Problem Information		
2644	2013				CT257	Implantable cardiac pacemakers	F1905	Device Revision or Replacement	A07	Electrical /Electronic Property Problem		
									A0708	Power Problem		
2645	2014				CT257	Implantable cardiac pacemakers	F1905	Device Revision or Replacement	A07	Electrical /Electronic Property Problem		
									A0708	Power Problem		
2646	2013		47270	Cardiac resynchronization therapy implantable defibrillator	CT2866	Implantable cardiac electrical stimulators			A26	Insufficient Device Problem Information		
2647	2014		47270	Cardiac resynchronization therapy implantable defibrillator	CT2866	Implantable cardiac electrical stimulators			A26	Insufficient Device Problem Information		
2648	2014		47270	Cardiac resynchronization therapy implantable defibrillator	CT2866	Implantable cardiac electrical stimulators			A26	Insufficient Device Problem Information		
2649	2013	All other reportable incidents			CT257	Implantable cardiac pacemakers	E2401	Insufficient Information	A13	Communication or Transmission Problem		
									A26	Insufficient Device Problem Information		
2650	2014				CT257	Implantable cardiac pacemakers	F1905	Device Revision or Replacement	A07	Electrical /Electronic Property Problem	29/08/2014	
									A0708	Power Problem	29/08/2014	
2651	2013		47270	Cardiac resynchronization therapy implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A07	Electrical /Electronic Property Problem		
									A0709	Device Sensing Problem		
2652	2013		47270	Cardiac resynchronization therapy implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	F1905	Device Revision or Replacement	A07	Electrical /Electronic Property Problem		
									A0708	Power Problem		
2653	2014				CT257	Implantable cardiac pacemakers	F1905	Device Revision or Replacement	A07	Electrical /Electronic Property Problem		
									A0708	Power Problem		
2654	2013		47270	Cardiac resynchronization therapy implantable defibrillator	CT2866	Implantable cardiac electrical stimulators			A26	Insufficient Device Problem Information		
2655	2013		47270	Cardiac resynchronization therapy implantable defibrillator	CT2866	Implantable cardiac electrical stimulators			A26	Insufficient Device Problem Information		
2656	2013		47270	Cardiac resynchronization therapy implantable defibrillator	CT2866	Implantable cardiac electrical stimulators			A26	Insufficient Device Problem Information		
2657	2013				CT257	Implantable cardiac pacemakers			A1104	Computer Operating System Problem		
2658	2013		47270	Cardiac resynchronization therapy implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A11	Computer Software Problem		
									A1102	Application Program Problem		
2659	2014		47270	Cardiac resynchronization therapy implantable defibrillator	CT2866	Implantable cardiac electrical stimulators			A26	Insufficient Device Problem Information		
2660	2014		47270	Cardiac resynchronization therapy implantable defibrillator	CT2866	Implantable cardiac electrical stimulators			A26	Insufficient Device Problem Information		
2661	2014		35852	Single-chamber implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	F1905	Device Revision or Replacement	A11	Computer Software Problem		
									A1102	Application Program Problem		
2662	2013	All other reportable incidents	47270	Cardiac resynchronization therapy implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	F1905	Device Revision or Replacement	A07	Electrical /Electronic Property Problem		
									A0708	Power Problem		
2663	2013	All other reportable incidents			CT257	Implantable cardiac pacemakers			A13	Communication or Transmission Problem		
									A26	Insufficient Device Problem Information		
2664	2013				CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A07	Electrical /Electronic Property Problem		
									A0709	Device Sensing Problem		
2665	2013				CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A11	Computer Software Problem		
									A1102	Application Program Problem		
2666	2013		47270	Cardiac resynchronization therapy implantable defibrillator	CT2866	Implantable cardiac electrical stimulators			A26	Insufficient Device Problem Information		
2667	2013				CT257	Implantable cardiac pacemakers	F1905	Device Revision or Replacement	A09	Output Problem		
									A0907	No Device Output		
2668	2014				CT257	Implantable cardiac pacemakers	F21	Unexpected Deterioration	A26	Insufficient Device Problem Information		
2669	2014				CT257	Implantable cardiac pacemakers	E0503	Embolism/Embolus	A09	Output Problem		
									A090404	Output Above Specifications		
2670	2014	All other reportable incidents			CT257	Implantable cardiac pacemakers	F1905	Device Revision or Replacement	A07	Electrical /Electronic Property Problem	24/04/2014	
									A0708	Power Problem	24/04/2014	
2671	2014		35852	Single-chamber implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E1906	Unspecified Infection	A01	Patient Device Interaction Problem		
									A17	Compatibility Problem		
							F1905	Device Revision or Replacement	A01	Patient Device Interaction Problem		
									A17	Compatibility Problem		
2672	2014				CT257	Implantable cardiac pacemakers			A26	Insufficient Device Problem Information		
2673	2013		47270	Cardiac resynchronization therapy implantable defibrillator	CT2866	Implantable cardiac electrical stimulators			A26	Insufficient Device Problem Information		
2674	2013		47270	Cardiac resynchronization therapy implantable defibrillator	CT2866	Implantable cardiac electrical stimulators			A26	Insufficient Device Problem Information		
2675	2013				CT257	Implantable cardiac pacemakers	F1905	Device Revision or Replacement	A07	Electrical /Electronic Property Problem		
									A0708	Power Problem		
2676	2014		35223	Endocardial/interventricular septal pacing lead	CT1684	Cardiac pacemakers and associated devices	E2008	Foreign Body In Patient	A04	Material Integrity Problem		
									A0404	Crack		
2677	2014		47270	Cardiac resynchronization therapy implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	F1905	Device Revision or Replacement	A07	Electrical /Electronic Property Problem		
									A0708	Power Problem		
2678	2013		47270	Cardiac resynchronization therapy implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	F1905	Device Revision or Replacement	A0104	Migration or Expulsion of Device		
									A07	Electrical /Electronic Property Problem		
									A0708	Power Problem		
2679	2014		35852	Single-chamber implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	F14	Prolonged Episode of Care	A01	Patient Device Interaction Problem		
									A17	Compatibility Problem		
2680	2014				CT257	Implantable cardiac pacemakers	F1905	Device Revision or Replacement	A04	Material Integrity Problem		
									A040103	Material Fragmentation		
2681	2013		47270	Cardiac resynchronization therapy implantable defibrillator	CT2866	Implantable cardiac electrical stimulators			A26	Insufficient Device Problem Information		

2682	2013				CT257	Implantable cardiac pacemakers	E2401	Insufficient Information	A26	Insufficient Device Problem Information		
2683	2013		47270	Cardiac resynchronization therapy implantable defibrillator	CT2866	Implantable cardiac electrical stimulators			A26	Insufficient Device Problem Information		
2684	2013				CT257	Implantable cardiac pacemakers			A07	Electrical /Electronic Property Problem		
									A0708	Power Problem		
2685	2013		47270	Cardiac resynchronization therapy implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A07	Electrical /Electronic Property Problem		
									A0709	Device Sensing Problem		
2686	2014		47270	Cardiac resynchronization therapy implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	F1905	Device Revision or Replacement	A11	Computer Software Problem	30/09/2014	
									A1102	Application Program Problem	30/09/2014	
2687	2014	All other reportable incidents			CT257	Implantable cardiac pacemakers	F1905	Device Revision or Replacement	A07	Electrical /Electronic Property Problem	19/09/2014	
									A0708	Power Problem	19/09/2014	
2688	2014				CT257	Implantable cardiac pacemakers	F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A07	Electrical /Electronic Property Problem		
									A0709	Device Sensing Problem		
2689	2014		47270	Cardiac resynchronization therapy implantable defibrillator	CT2866	Implantable cardiac electrical stimulators			A26	Insufficient Device Problem Information		
2690	2014		47270	Cardiac resynchronization therapy implantable defibrillator	CT2866	Implantable cardiac electrical stimulators			A26	Insufficient Device Problem Information		
2691	2014		47270	Cardiac resynchronization therapy implantable defibrillator	CT2866	Implantable cardiac electrical stimulators			A26	Insufficient Device Problem Information		
2692	2014				CT257	Implantable cardiac pacemakers	F21	Unexpected Deterioration	A07	Electrical /Electronic Property Problem		
									A0709	Device Sensing Problem		
2693	2013	All other reportable incidents	47270	Cardiac resynchronization therapy implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	F1905	Device Revision or Replacement	A07	Electrical /Electronic Property Problem	12/02/2013	
									A0708	Power Problem	12/02/2013	
2694	2013	All other reportable incidents	47270	Cardiac resynchronization therapy implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	F1905	Device Revision or Replacement	A07	Electrical /Electronic Property Problem	11/07/2013	
									A0708	Power Problem	11/07/2013	
2695	2013				CT257	Implantable cardiac pacemakers	F21	Unexpected Deterioration	A13	Communication or Transmission Problem		
									A26	Insufficient Device Problem Information		
2696	2013		47270	Cardiac resynchronization therapy implantable defibrillator	CT2866	Implantable cardiac electrical stimulators			A26	Insufficient Device Problem Information		
2697	2013		47270	Cardiac resynchronization therapy implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	F1905	Device Revision or Replacement	A13	Communication or Transmission Problem		
									A26	Insufficient Device Problem Information		
2698	2013				CT257	Implantable cardiac pacemakers	F1905	Device Revision or Replacement	A07	Electrical /Electronic Property Problem		
									A0708	Power Problem		
									A0709	Device Sensing Problem		
2699	2013	All other reportable incidents	47270	Cardiac resynchronization therapy implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	F1905	Device Revision or Replacement	A07	Electrical /Electronic Property Problem	22/01/2013	
									A0708	Power Problem	22/01/2013	
2700	2014	All other reportable incidents			CT257	Implantable cardiac pacemakers	F1905	Device Revision or Replacement	A07	Electrical /Electronic Property Problem	26/03/2014	
									A0708	Power Problem	26/03/2014	
2701	2013	All other reportable incidents			CT257	Implantable cardiac pacemakers	F1905	Device Revision or Replacement	A13	Communication or Transmission Problem	24/06/2013	
									A26	Insufficient Device Problem Information	24/06/2013	
2702	2013		47270	Cardiac resynchronization therapy implantable defibrillator	CT2866	Implantable cardiac electrical stimulators			A26	Insufficient Device Problem Information		
2703	2013		35852	Single-chamber implantable defibrillator	CT2866	Implantable cardiac electrical stimulators			A07	Electrical /Electronic Property Problem		
									A0709	Device Sensing Problem		
2704	2013		47270	Cardiac resynchronization therapy implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	F14	Prolonged Episode of Care	A01	Patient Device Interaction Problem		
									A17	Compatibility Problem		
2705	2014				CT257	Implantable cardiac pacemakers	E2401	Insufficient Information	A26	Insufficient Device Problem Information		
2706	2014		47270	Cardiac resynchronization therapy implantable defibrillator	CT2866	Implantable cardiac electrical stimulators			A26	Insufficient Device Problem Information		
2707	2013		47270	Cardiac resynchronization therapy implantable defibrillator	CT2866	Implantable cardiac electrical stimulators			A26	Insufficient Device Problem Information		
2708	2013		47270	Cardiac resynchronization therapy implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E0402	Hypersensitivity/Allergic Reaction	A26	Insufficient Device Problem Information		
2709	2013		47270	Cardiac resynchronization therapy implantable defibrillator	CT2866	Implantable cardiac electrical stimulators			A26	Insufficient Device Problem Information		
2710	2014				CT257	Implantable cardiac pacemakers	E0402	Hypersensitivity/Allergic Reaction	A26	Insufficient Device Problem Information		
2711	2013		47270	Cardiac resynchronization therapy implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A26	Insufficient Device Problem Information		
2712	2013				CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A26	Insufficient Device Problem Information		
2713	2013		47270	Cardiac resynchronization therapy implantable defibrillator	CT2866	Implantable cardiac electrical stimulators			A26	Insufficient Device Problem Information		
2714	2013		47270	Cardiac resynchronization therapy implantable defibrillator	CT2866	Implantable cardiac electrical stimulators			A26	Insufficient Device Problem Information		
2715	2013		47270	Cardiac resynchronization therapy implantable defibrillator	CT2866	Implantable cardiac electrical stimulators			A26	Insufficient Device Problem Information		
2716	2013		47270	Cardiac resynchronization therapy implantable defibrillator	CT2866	Implantable cardiac electrical stimulators			A26	Insufficient Device Problem Information		
2717	2014		47270	Cardiac resynchronization therapy implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	F1905	Device Revision or Replacement	A07	Electrical /Electronic Property Problem		
									A0708	Power Problem		
2718	2014				CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A09	Output Problem		
									A0907	No Device Output		
2719	2014				CT257	Implantable cardiac pacemakers	E0402	Hypersensitivity/Allergic Reaction	A26	Insufficient Device Problem Information		
2720	2014				CT257	Implantable cardiac pacemakers	F1905	Device Revision or Replacement	A07	Electrical /Electronic Property Problem		
									A0708	Power Problem		
2721	2014				CT257	Implantable cardiac pacemakers	F1905	Device Revision or Replacement	A07	Electrical /Electronic Property Problem		
									A0708	Power Problem		
2722	2013				CT257	Implantable cardiac pacemakers	F1905	Device Revision or Replacement	A07	Electrical /Electronic Property Problem		
									A0709	Device Sensing Problem		
2723	2013		47270	Cardiac resynchronization therapy implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A07	Electrical /Electronic Property Problem		
									A0709	Device Sensing Problem		
2724	2014		47270	Cardiac resynchronization therapy implantable defibrillator	CT2866	Implantable cardiac electrical stimulators			A26	Insufficient Device Problem Information		
2725	2014		47270	Cardiac resynchronization therapy implantable defibrillator	CT2866	Implantable cardiac electrical stimulators			A26	Insufficient Device Problem Information		
2726	2013	All other reportable incidents	47270	Cardiac resynchronization therapy implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	F1905	Device Revision or Replacement	A07	Electrical /Electronic Property Problem	18/10/2012	
									A0708	Power Problem	18/10/2012	
2727	2013	All other reportable incidents			CT257	Implantable cardiac pacemakers	F1905	Device Revision or Replacement	A07	Electrical /Electronic Property Problem	16/11/2012	
									A0708	Power Problem	16/11/2012	
2728	2014		47270	Cardiac resynchronization therapy implantable defibrillator	CT2866	Implantable cardiac electrical stimulators			A26	Insufficient Device Problem Information		
2729	2014	Death			CT257	Implantable cardiac pacemakers	F05	Delay to Treatment/ Therapy	A07	Electrical /Electronic Property Problem	03/05/2014	
									A0709	Device Sensing Problem	03/05/2014	

									A09	Output Problem	03/05/2014	
									A090404	Output Above Specifications	03/05/2014	
2730	2013		47270	Cardiac resynchronization therapy implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E2401	Insufficient Information	A26	Insufficient Device Problem Information		
2731	2013		47270	Cardiac resynchronization therapy implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E2401	Insufficient Information	A26	Insufficient Device Problem Information		
2732	2013		47270	Cardiac resynchronization therapy implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E0402	Hypersensitivity/Allergic Reaction	A26	Insufficient Device Problem Information		
2733	2014		47270	Cardiac resynchronization therapy implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	F1905	Device Revision or Replacement	A07	Electrical /Electronic Property Problem		
									A0708	Power Problem		
2734	2014		47270	Cardiac resynchronization therapy implantable defibrillator	CT2866	Implantable cardiac electrical stimulators			A26	Insufficient Device Problem Information		
2735	2014				CT257	Implantable cardiac pacemakers	F1905	Device Revision or Replacement	A07	Electrical /Electronic Property Problem	14/05/2014	
									A0708	Power Problem	14/05/2014	
2736	2014				CT257	Implantable cardiac pacemakers	F1905	Device Revision or Replacement	A07	Electrical /Electronic Property Problem	12/05/2014	
									A0708	Power Problem	12/05/2014	
2737	2014				CT257	Implantable cardiac pacemakers			A26	Insufficient Device Problem Information		
2738	2013				CT257	Implantable cardiac pacemakers	E2401	Insufficient Information	A07	Electrical /Electronic Property Problem		
									A0708	Power Problem		
2739	2014	All other reportable incidents			CT257	Implantable cardiac pacemakers	E2401	Insufficient Information	A13	Communication or Transmission Problem	10/09/2014	
									A26	Insufficient Device Problem Information	10/09/2014	
2740	2013		47270	Cardiac resynchronization therapy implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	F15	Recognized Device or Procedural Complication	A26	Insufficient Device Problem Information		
2741	2014		47270	Cardiac resynchronization therapy implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E0402	Hypersensitivity/Allergic Reaction	A26	Insufficient Device Problem Information		
2742	2013				CT257	Implantable cardiac pacemakers	F1905	Device Revision or Replacement	A07	Electrical /Electronic Property Problem	28/02/2013	
									A0708	Power Problem	28/02/2013	
2743	2014	All other reportable incidents			CT257	Implantable cardiac pacemakers	F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A09	Output Problem		
									A090404	Output Above Specifications		
							F21	Unexpected Deterioration	A09	Output Problem		
									A090404	Output Above Specifications		
2744	2013		47270	Cardiac resynchronization therapy implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A26	Insufficient Device Problem Information		
2745	2013		47270	Cardiac resynchronization therapy implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	F15	Recognized Device or Procedural Complication	A26	Insufficient Device Problem Information		
2746	2013		47270	Cardiac resynchronization therapy implantable defibrillator	CT2866	Implantable cardiac electrical stimulators			A09	Output Problem		
									A0907	No Device Output		
2747	2013		47270	Cardiac resynchronization therapy implantable defibrillator	CT2866	Implantable cardiac electrical stimulators			A26	Insufficient Device Problem Information		
2748	2014	All other reportable incidents	47270	Cardiac resynchronization therapy implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A07	Electrical /Electronic Property Problem		
									A0708	Power Problem		
2749	2014				CT257	Implantable cardiac pacemakers	E2401	Insufficient Information	A12	Connection Problem		
									A1208	Fitting Problem		
2750	2013	All other reportable incidents			CT1684	Cardiac pacemakers and associated devices	F15	Recognized Device or Procedural Complication	A23	Use of Device Problem		
2751	2014	All other reportable incidents	47270	Cardiac resynchronization therapy implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	F1905	Device Revision or Replacement	A07	Electrical /Electronic Property Problem	28/04/2014	
									A0708	Power Problem	28/04/2014	
2752	2013				CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A07	Electrical /Electronic Property Problem		
									A0708	Power Problem		
2753	2014		35852	Single-chamber implantable defibrillator	CT2866	Implantable cardiac electrical stimulators			A1304	Telemetry Discrepancy		
									A26	Insufficient Device Problem Information		
2754	2013	All other reportable incidents			CT257	Implantable cardiac pacemakers	F1905	Device Revision or Replacement	A07	Electrical /Electronic Property Problem		
									A0708	Power Problem		
2755	2014	All other reportable incidents			CT257	Implantable cardiac pacemakers	F1905	Device Revision or Replacement	A07	Electrical /Electronic Property Problem	16/05/2014	
									A0708	Power Problem	16/05/2014	
2756	2014		47270	Cardiac resynchronization therapy implantable defibrillator	CT2866	Implantable cardiac electrical stimulators			A26	Insufficient Device Problem Information		
2757	2013		47270	Cardiac resynchronization therapy implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E0402	Hypersensitivity/Allergic Reaction	A26	Insufficient Device Problem Information		
2758	2013		47270	Cardiac resynchronization therapy implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	F21	Unexpected Deterioration	A26	Insufficient Device Problem Information		
2759	2013		47270	Cardiac resynchronization therapy implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	F21	Unexpected Deterioration	A26	Insufficient Device Problem Information		
2760	2014		47270	Cardiac resynchronization therapy implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	F1905	Device Revision or Replacement	A07	Electrical /Electronic Property Problem		
									A0708	Power Problem		
2761	2013		47270	Cardiac resynchronization therapy implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	F21	Unexpected Deterioration	A26	Insufficient Device Problem Information		
2762	2013				CT257	Implantable cardiac pacemakers	F1905	Device Revision or Replacement	A07	Electrical /Electronic Property Problem		
									A0709	Device Sensing Problem		
2763	2013		47270	Cardiac resynchronization therapy implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E0402	Hypersensitivity/Allergic Reaction	A26	Insufficient Device Problem Information		
2764	2014		47270	Cardiac resynchronization therapy implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E2401	Insufficient Information	A07	Electrical /Electronic Property Problem		
									A0708	Power Problem		
2765	2013				CT257	Implantable cardiac pacemakers	E0139	Unspecified Nervous System Problem	A09	Output Problem		
									A090404	Output Above Specifications		
2766	2013		47270	Cardiac resynchronization therapy implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E2401	Insufficient Information	A26	Insufficient Device Problem Information		
2767	2014		47270	Cardiac resynchronization therapy implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	F1905	Device Revision or Replacement	A07	Electrical /Electronic Property Problem		
									A0708	Power Problem		
2768	2013		47270	Cardiac resynchronization therapy implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	F15	Recognized Device or Procedural Complication	A26	Insufficient Device Problem Information		
2769	2013				CT257	Implantable cardiac pacemakers	F1905	Device Revision or Replacement	A07	Electrical /Electronic Property Problem		
									A0708	Power Problem		
2770	2013				CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A07	Electrical /Electronic Property Problem		
									A0708	Power Problem		
2771	2013				CT257	Implantable cardiac pacemakers			A07	Electrical /Electronic Property Problem		
									A0708	Power Problem		
2772	2014				CT257	Implantable cardiac pacemakers			A26	Insufficient Device Problem Information		
2773	2014		47270	Cardiac resynchronization therapy implantable defibrillator	CT2866	Implantable cardiac electrical stimulators			A26	Insufficient Device Problem Information		
2774	2013		47270	Cardiac resynchronization therapy implantable defibrillator	CT2866	Implantable cardiac electrical stimulators			A26	Insufficient Device Problem Information		
2775	2013	All other reportable incidents			CT257	Implantable cardiac pacemakers	F1905	Device Revision or Replacement	A07	Electrical /Electronic Property Problem	28/03/2013	

									A0709	Device Sensing Problem	28/03/2013	
2776	2014		47270	Cardiac resynchronization therapy implantable defibrillator	CT2866	Implantable cardiac electrical stimulators			A26	Insufficient Device Problem Information		
2777	2013		47270	Cardiac resynchronization therapy implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E2401	Insufficient Information	A26	Insufficient Device Problem Information		
2778	2014		47270	Cardiac resynchronization therapy implantable defibrillator	CT2866	Implantable cardiac electrical stimulators			A26	Insufficient Device Problem Information		
2779	2014		47270	Cardiac resynchronization therapy implantable defibrillator	CT2866	Implantable cardiac electrical stimulators			A26	Insufficient Device Problem Information		
2780	2014				CT257	Implantable cardiac pacemakers			A26	Insufficient Device Problem Information		
2781	2013				CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A07	Electrical /Electronic Property Problem		
									A0708	Power Problem		
2782	2014				CT257	Implantable cardiac pacemakers	E2401	Insufficient Information	A13	Communication or Transmission Problem		
									A26	Insufficient Device Problem Information		
2783	2014				CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A11	Computer Software Problem		
									A1102	Application Program Problem		
									A15	Activation, Positioning or Separation Problem		
									A150102	Difficult or Delayed Activation		
2784	2013				CT257	Implantable cardiac pacemakers	E2401	Insufficient Information	A07	Electrical /Electronic Property Problem		
									A0709	Device Sensing Problem		
2785	2013	All other reportable incidents	47270	Cardiac resynchronization therapy implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	F1905	Device Revision or Replacement	A07	Electrical /Electronic Property Problem	02/10/2012	
									A0708	Power Problem	02/10/2012	
2786	2013				CT257	Implantable cardiac pacemakers	E2401	Insufficient Information	A07	Electrical /Electronic Property Problem		
									A0709	Device Sensing Problem		
2787	2013				CT257	Implantable cardiac pacemakers			A07	Electrical /Electronic Property Problem		
									A0709	Device Sensing Problem		
2788	2013		47270	Cardiac resynchronization therapy implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A07	Electrical /Electronic Property Problem		
									A0709	Device Sensing Problem		
2789	2014	All other reportable incidents	47270	Cardiac resynchronization therapy implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	F1905	Device Revision or Replacement	A07	Electrical /Electronic Property Problem	23/05/2014	
									A0708	Power Problem	23/05/2014	
2790	2014		47270	Cardiac resynchronization therapy implantable defibrillator	CT2866	Implantable cardiac electrical stimulators			A26	Insufficient Device Problem Information		
2791	2014		47270	Cardiac resynchronization therapy implantable defibrillator	CT2866	Implantable cardiac electrical stimulators			A26	Insufficient Device Problem Information		
2792	2013		47270	Cardiac resynchronization therapy implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	F1905	Device Revision or Replacement	A07	Electrical /Electronic Property Problem		
									A0709	Device Sensing Problem		
2793	2014		47270	Cardiac resynchronization therapy implantable defibrillator	CT2866	Implantable cardiac electrical stimulators			A26	Insufficient Device Problem Information		
2794	2014				CT257	Implantable cardiac pacemakers	E2401	Insufficient Information	A07	Electrical /Electronic Property Problem		
									A0708	Power Problem		
2795	2013				CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A07	Electrical /Electronic Property Problem		
									A0708	Power Problem		
									A13	Communication or Transmission Problem		
									A26	Insufficient Device Problem Information		
2796	2013				CT257	Implantable cardiac pacemakers	F1905	Device Revision or Replacement	A07	Electrical /Electronic Property Problem		
									A0709	Device Sensing Problem		
2797	2014				CT257	Implantable cardiac pacemakers	F1905	Device Revision or Replacement	A07	Electrical /Electronic Property Problem		
									A0708	Power Problem		
2798	2013		47270	Cardiac resynchronization therapy implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	F15	Recognized Device or Procedural Complication	A26	Insufficient Device Problem Information		
2799	2013				CT257	Implantable cardiac pacemakers			A26	Insufficient Device Problem Information		
2800	2013		47270	Cardiac resynchronization therapy implantable defibrillator	CT2866	Implantable cardiac electrical stimulators			A26	Insufficient Device Problem Information		
2801	2013				CT257	Implantable cardiac pacemakers	E2401	Insufficient Information	A07	Electrical /Electronic Property Problem		
									A0708	Power Problem		
2802	2014		47270	Cardiac resynchronization therapy implantable defibrillator	CT2866	Implantable cardiac electrical stimulators			A26	Insufficient Device Problem Information		
2803	2014		47270	Cardiac resynchronization therapy implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	F1905	Device Revision or Replacement	A07	Electrical /Electronic Property Problem		
									A0708	Power Problem		
2804	2014		47270	Cardiac resynchronization therapy implantable defibrillator	CT2866	Implantable cardiac electrical stimulators			A26	Insufficient Device Problem Information		
2805	2014		47270	Cardiac resynchronization therapy implantable defibrillator	CT2866	Implantable cardiac electrical stimulators			A26	Insufficient Device Problem Information		
2806	2013		47270	Cardiac resynchronization therapy implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	F15	Recognized Device or Procedural Complication	A26	Insufficient Device Problem Information		
2807	2013				CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A1304	Telemetry Discrepancy		
									A26	Insufficient Device Problem Information		
2808	2014				CT257	Implantable cardiac pacemakers			A26	Insufficient Device Problem Information		
2809	2014	All other reportable incidents	47270	Cardiac resynchronization therapy implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A07	Electrical /Electronic Property Problem		
									A0709	Device Sensing Problem		
2810	2013		47270	Cardiac resynchronization therapy implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A26	Insufficient Device Problem Information		
2811	2013		47270	Cardiac resynchronization therapy implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E2401	Insufficient Information	A26	Insufficient Device Problem Information		
2812	2013		47270	Cardiac resynchronization therapy implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E0402	Hypersensitivity/Allergic Reaction	A26	Insufficient Device Problem Information		
2813	2013				CT257	Implantable cardiac pacemakers	F1905	Device Revision or Replacement	A13	Communication or Transmission Problem		
									A26	Insufficient Device Problem Information		
2814	2013				CT257	Implantable cardiac pacemakers	F1905	Device Revision or Replacement	A05	Mechanical Problem		
									A0501	Detachment of Device or Device Component		
									A07	Electrical /Electronic Property Problem		
									A0709	Device Sensing Problem		
2815	2014		35852	Single-chamber implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E2104	Electric Shock	A09	Output Problem		
									A090404	Output Above Specifications		
2816	2013	All other reportable incidents			CT1684	Cardiac pacemakers and associated devices	F05	Delay to Treatment/ Therapy	A09	Output Problem		
									A0907	No Device Output		
2817	2014		47270	Cardiac resynchronization therapy implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E0402	Hypersensitivity/Allergic Reaction	A26	Insufficient Device Problem Information		
2818	2013	All other reportable incidents			CT257	Implantable cardiac pacemakers	F1905	Device Revision or Replacement	A07	Electrical /Electronic Property Problem	27/06/2013	
									A0708	Power Problem	27/06/2013	
2819	2013	All other reportable incidents			CT257	Implantable cardiac pacemakers	F14	Prolonged Episode of Care	A07	Electrical /Electronic Property Problem		
									A0708	Power Problem		

2820	2014		47270	Cardiac resynchronization therapy implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A07	Electrical /Electronic Property Problem		
									A0709	Device Sensing Problem		
2821	2014		47270	Cardiac resynchronization therapy implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A13	Communication or Transmission Problem		
									A26	Insufficient Device Problem Information		
2822	2013	All other reportable incidents			CT257	Implantable cardiac pacemakers	F1905	Device Revision or Replacement	A07	Electrical /Electronic Property Problem	11/02/2013	
									A0708	Power Problem	11/02/2013	
2823	2013				CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A07	Electrical /Electronic Property Problem		
									A0708	Power Problem		
2824	2013				CT257	Implantable cardiac pacemakers			A1304	Telemetry Discrepancy		
									A26	Insufficient Device Problem Information		
2825	2013				CT257	Implantable cardiac pacemakers	F1905	Device Revision or Replacement	A07	Electrical /Electronic Property Problem		
									A0709	Device Sensing Problem		
2826	2014		47270	Cardiac resynchronization therapy implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E2401	Insufficient Information	A26	Insufficient Device Problem Information		
2827	2014		47270	Cardiac resynchronization therapy implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	F1905	Device Revision or Replacement	A07	Electrical /Electronic Property Problem		
									A0708	Power Problem		
2828	2014				CT257	Implantable cardiac pacemakers			A26	Insufficient Device Problem Information		
2829	2014		47270	Cardiac resynchronization therapy implantable defibrillator	CT2866	Implantable cardiac electrical stimulators			A26	Insufficient Device Problem Information		
2830	2014	All other reportable incidents			CT257	Implantable cardiac pacemakers			A07	Electrical /Electronic Property Problem		
									A0708	Power Problem		
2831	2013				CT257	Implantable cardiac pacemakers			A07	Electrical /Electronic Property Problem		
									A0708	Power Problem		
2832	2014		47270	Cardiac resynchronization therapy implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E2109	Implant Pain	A0104	Migration or Expulsion of Device		
									A07	Electrical /Electronic Property Problem		
									A0709	Device Sensing Problem		
							F14	Prolonged Episode of Care	A0104	Migration or Expulsion of Device		
									A07	Electrical /Electronic Property Problem		
									A0709	Device Sensing Problem		
2833	2014				CT257	Implantable cardiac pacemakers	E2401	Insufficient Information	A11	Computer Software Problem		
									A1102	Application Program Problem		
2834	2013				CT257	Implantable cardiac pacemakers	E0402	Hypersensitivity/Allergic Reaction	A07	Electrical /Electronic Property Problem		
									A0709	Device Sensing Problem		
2835	2013				CT257	Implantable cardiac pacemakers	E2401	Insufficient Information	A07	Electrical /Electronic Property Problem		
									A0708	Power Problem		
2836	2013		47270	Cardiac resynchronization therapy implantable defibrillator	CT2866	Implantable cardiac electrical stimulators			A26	Insufficient Device Problem Information		
2837	2013	All other reportable incidents			CT257	Implantable cardiac pacemakers	F1905	Device Revision or Replacement	A07	Electrical /Electronic Property Problem	17/07/2013	
									A0709	Device Sensing Problem	17/07/2013	
2838	2014		47270	Cardiac resynchronization therapy implantable defibrillator	CT2866	Implantable cardiac electrical stimulators			A26	Insufficient Device Problem Information		
2839	2014	All other reportable incidents			CT257	Implantable cardiac pacemakers	E2401	Insufficient Information	A07	Electrical /Electronic Property Problem		
									A0708	Power Problem		
2840	2013	All other reportable incidents			CT257	Implantable cardiac pacemakers	F1905	Device Revision or Replacement	A07	Electrical /Electronic Property Problem		
									A0708	Power Problem		
2841	2014		47270	Cardiac resynchronization therapy implantable defibrillator	CT2866	Implantable cardiac electrical stimulators			A26	Insufficient Device Problem Information		
2842	2014		47270	Cardiac resynchronization therapy implantable defibrillator	CT2866	Implantable cardiac electrical stimulators			A26	Insufficient Device Problem Information		
2843	2013		47270	Cardiac resynchronization therapy implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E2401	Insufficient Information	A26	Insufficient Device Problem Information		
2844	2014	All other reportable incidents			CT257	Implantable cardiac pacemakers	F1905	Device Revision or Replacement	A07	Electrical /Electronic Property Problem	10/10/2014	
									A0708	Power Problem	10/10/2014	
2845	2013		47270	Cardiac resynchronization therapy implantable defibrillator	CT2866	Implantable cardiac electrical stimulators			A26	Insufficient Device Problem Information		
2846	2013		47270	Cardiac resynchronization therapy implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A26	Insufficient Device Problem Information		
2847	2014		47270	Cardiac resynchronization therapy implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	F1905	Device Revision or Replacement	A07	Electrical /Electronic Property Problem		
									A0708	Power Problem		
2848	2014				CT257	Implantable cardiac pacemakers	F1905	Device Revision or Replacement	A07	Electrical /Electronic Property Problem		
									A0708	Power Problem		
2849	2013		47270	Cardiac resynchronization therapy implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A09	Output Problem		
									A090404	Output Above Specifications		
2850	2013		47270	Cardiac resynchronization therapy implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	F15	Recognized Device or Procedural Complication	A26	Insufficient Device Problem Information		
2851	2013		47270	Cardiac resynchronization therapy implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	F21	Unexpected Deterioration	A26	Insufficient Device Problem Information		
2852	2014				CT257	Implantable cardiac pacemakers	F1905	Device Revision or Replacement	A0104	Migration or Expulsion of Device		
2853	2014		47270	Cardiac resynchronization therapy implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	F1905	Device Revision or Replacement	A07	Electrical /Electronic Property Problem		
									A0708	Power Problem		
2854	2014				CT257	Implantable cardiac pacemakers	F1905	Device Revision or Replacement	A07	Electrical /Electronic Property Problem		
									A0708	Power Problem		
2855	2014				CT257	Implantable cardiac pacemakers	F1905	Device Revision or Replacement	A07	Electrical /Electronic Property Problem		
									A0708	Power Problem		
2856	2014				CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A17	Compatibility Problem		
									A1702	Device-Device Incompatibility		
									A23	Use of Device Problem		
2857	2013		47270	Cardiac resynchronization therapy implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	F1905	Device Revision or Replacement	A07	Electrical /Electronic Property Problem		
									A0708	Power Problem		
2858	2013		47270	Cardiac resynchronization therapy implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	F1905	Device Revision or Replacement	A07	Electrical /Electronic Property Problem		
									A0708	Power Problem		
2859	2014		47270	Cardiac resynchronization therapy implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A07	Electrical /Electronic Property Problem		
									A0709	Device Sensing Problem		
2860	2014		47270	Cardiac resynchronization therapy implantable defibrillator	CT2866	Implantable cardiac electrical stimulators			A26	Insufficient Device Problem Information		
2861	2014		47270	Cardiac resynchronization therapy implantable defibrillator	CT2866	Implantable cardiac electrical stimulators			A26	Insufficient Device Problem Information		
2862	2014				CT257	Implantable cardiac pacemakers			A26	Insufficient Device Problem Information		
2863	2013		47270	Cardiac resynchronization therapy implantable defibrillator	CT2866	Implantable cardiac electrical stimulators			A26	Insufficient Device Problem Information		
2864	2014		47270	Cardiac resynchronization therapy implantable	CT2866	Implantable cardiac electrical stimulators	F1905	Device Revision or Replacement	A07	Electrical /Electronic Property Problem	24/10/2014	

				defibrillator					A0708	Power Problem	24/10/2014	
2865	2013				CT257	Implantable cardiac pacemakers	E0402	Hypersensitivity/Allergic Reaction	A26	Insufficient Device Problem Information		
2866	2014				CT257	Implantable cardiac pacemakers	E2401	Insufficient Information	A07	Electrical /Electronic Property Problem		
									A0708	Power Problem		
2867	2014		47270	Cardiac resynchronization therapy implantable defibrillator	CT2866	Implantable cardiac electrical stimulators			A26	Insufficient Device Problem Information		
2868	2014		47270	Cardiac resynchronization therapy implantable defibrillator	CT2866	Implantable cardiac electrical stimulators			A26	Insufficient Device Problem Information		
2869	2014		47270	Cardiac resynchronization therapy implantable defibrillator	CT2866	Implantable cardiac electrical stimulators			A26	Insufficient Device Problem Information		
2870	2014	All other reportable incidents			CT257	Implantable cardiac pacemakers	F1905	Device Revision or Replacement	A07	Electrical /Electronic Property Problem		
									A0708	Power Problem		
2871	2014				CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A07	Electrical /Electronic Property Problem		
									A0708	Power Problem		
2872	2014				CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A07	Electrical /Electronic Property Problem		
									A0708	Power Problem		
2873	2014				CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A07	Electrical /Electronic Property Problem		
									A0708	Power Problem		
2874	2014	All other reportable incidents			CT257	Implantable cardiac pacemakers	F1905	Device Revision or Replacement	A07	Electrical /Electronic Property Problem		
									A0708	Power Problem		
2875	2013				CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A21	Labeling, Instructions for Use or Training Problem		
2876	2014				CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A1501	Activation Problem		
									A23	Use of Device Problem		
2877	2014				CT257	Implantable cardiac pacemakers	F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A07	Electrical /Electronic Property Problem		
									A0709	Device Sensing Problem		
2878	2014				CT257	Implantable cardiac pacemakers	F1905	Device Revision or Replacement	A07	Electrical /Electronic Property Problem		
									A0708	Power Problem		
2879	2014				CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A07	Electrical /Electronic Property Problem		
									A0708	Power Problem		
2880	2014				CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A07	Electrical /Electronic Property Problem		
									A0708	Power Problem		
2881	2013		47270	Cardiac resynchronization therapy implantable defibrillator	CT2866	Implantable cardiac electrical stimulators			A26	Insufficient Device Problem Information		
2882	2013		47270	Cardiac resynchronization therapy implantable defibrillator	CT2866	Implantable cardiac electrical stimulators			A26	Insufficient Device Problem Information		
2883	2014				CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A11	Computer Software Problem		
									A1102	Application Program Problem		
2884	2014		47270	Cardiac resynchronization therapy implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A07	Electrical /Electronic Property Problem		
									A0709	Device Sensing Problem		
2885	2014		47270	Cardiac resynchronization therapy implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E0402	Hypersensitivity/Allergic Reaction	A26	Insufficient Device Problem Information		
2886	2014				CT257	Implantable cardiac pacemakers	E0402	Hypersensitivity/Allergic Reaction	A26	Insufficient Device Problem Information		
2887	2013	All other reportable incidents			CT257	Implantable cardiac pacemakers	F1905	Device Revision or Replacement	A07	Electrical /Electronic Property Problem		
									A0708	Power Problem		
2888	2013	All other reportable incidents			CT257	Implantable cardiac pacemakers			A07	Electrical /Electronic Property Problem		
									A0708	Power Problem		
2889	2014				CT257	Implantable cardiac pacemakers	E2401	Insufficient Information	A07	Electrical /Electronic Property Problem		
									A0708	Power Problem		
2890	2014	All other reportable incidents			CT257	Implantable cardiac pacemakers	F1905	Device Revision or Replacement	A07	Electrical /Electronic Property Problem	02/01/2014	
									A0708	Power Problem	02/01/2014	
2891	2013		47270	Cardiac resynchronization therapy implantable defibrillator	CT2866	Implantable cardiac electrical stimulators			A26	Insufficient Device Problem Information		
2892	2014	All other reportable incidents	47270	Cardiac resynchronization therapy implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	F1905	Device Revision or Replacement	A07	Electrical /Electronic Property Problem		
									A0708	Power Problem		
2893	2014		47270	Cardiac resynchronization therapy implantable defibrillator	CT2866	Implantable cardiac electrical stimulators			A26	Insufficient Device Problem Information		
2894	2014		47270	Cardiac resynchronization therapy implantable defibrillator	CT2866	Implantable cardiac electrical stimulators			A26	Insufficient Device Problem Information		
2895	2014		47270	Cardiac resynchronization therapy implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E0402	Hypersensitivity/Allergic Reaction	A26	Insufficient Device Problem Information		
2896	2014		47270	Cardiac resynchronization therapy implantable defibrillator	CT2866	Implantable cardiac electrical stimulators			A26	Insufficient Device Problem Information		
2897	2013	All other reportable incidents			CT257	Implantable cardiac pacemakers	E2401	Insufficient Information	A07	Electrical /Electronic Property Problem		
									A0708	Power Problem		
2898	2013	All other reportable incidents			CT257	Implantable cardiac pacemakers			A07	Electrical /Electronic Property Problem		
									A0708	Power Problem		
2899	2014				CT257	Implantable cardiac pacemakers	F1905	Device Revision or Replacement	A07	Electrical /Electronic Property Problem		
									A0708	Power Problem		
2900	2013	All other reportable incidents			CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A1501	Activation Problem		
									A23	Use of Device Problem		
2901	2014		47270	Cardiac resynchronization therapy implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	F21	Unexpected Deterioration	A26	Insufficient Device Problem Information		
2902	2014				CT257	Implantable cardiac pacemakers	F1905	Device Revision or Replacement	A07	Electrical /Electronic Property Problem		
									A0708	Power Problem		
2903	2014		47270	Cardiac resynchronization therapy implantable defibrillator	CT2866	Implantable cardiac electrical stimulators			A26	Insufficient Device Problem Information		
2904	2014				CT257	Implantable cardiac pacemakers			A26	Insufficient Device Problem Information		
2905	2014	All other reportable incidents	47270	Cardiac resynchronization therapy implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E2401	Insufficient Information	A07	Electrical /Electronic Property Problem	14/10/2014	
									A0708	Power Problem	14/10/2014	
2906	2013		47270	Cardiac resynchronization therapy implantable defibrillator	CT2866	Implantable cardiac electrical stimulators			A26	Insufficient Device Problem Information		
2907	2013		47270	Cardiac resynchronization therapy implantable defibrillator	CT2866	Implantable cardiac electrical stimulators			A26	Insufficient Device Problem Information		
2908	2014				CT257	Implantable cardiac pacemakers	F1905	Device Revision or Replacement	A07	Electrical /Electronic Property Problem		
									A0708	Power Problem		
2909	2013	All other reportable incidents			CT257	Implantable cardiac pacemakers	F1905	Device Revision or Replacement	A07	Electrical /Electronic Property Problem	15/03/2013	
									A0708	Power Problem	15/03/2013	
2910	2013		47270	Cardiac resynchronization therapy implantable defibrillator	CT2866	Implantable cardiac electrical stimulators			A26	Insufficient Device Problem Information		
2911	2014		47270	Cardiac resynchronization therapy implantable defibrillator	CT2866	Implantable cardiac electrical stimulators			A26	Insufficient Device Problem Information		

2912	2014	All other reportable incidents			CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A09	Output Problem		
									A0907	No Device Output		
2913	2013		47270	Cardiac resynchronization therapy implantable defibrillator	CT2866	Implantable cardiac electrical stimulators			A09	Output Problem		
									A090404	Output Above Specifications		
2914	2014		47270	Cardiac resynchronization therapy implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	F1905	Device Revision or Replacement	A13	Communication or Transmission Problem		
									A1304	Telemetry Discrepancy		
									A26	Insufficient Device Problem Information		
2915	2013				CT257	Implantable cardiac pacemakers	F15	Recognized Device or Procedural Complication	A26	Insufficient Device Problem Information		
2916	2014				CT257	Implantable cardiac pacemakers	F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A07	Electrical /Electronic Property Problem		
									A0709	Device Sensing Problem		
2917	2014				CT257	Implantable cardiac pacemakers	E0402	Hypersensitivity/Allergic Reaction	A26	Insufficient Device Problem Information		
2918	2014		47270	Cardiac resynchronization therapy implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E2401	Insufficient Information	A07	Electrical /Electronic Property Problem		
									A0709	Device Sensing Problem		
2919	2013	All other reportable incidents			CT257	Implantable cardiac pacemakers	F1905	Device Revision or Replacement	A07	Electrical /Electronic Property Problem	14/03/2013	
									A0708	Power Problem	14/03/2013	
2920	2014		47270	Cardiac resynchronization therapy implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	F1905	Device Revision or Replacement	A07	Electrical /Electronic Property Problem		
									A0708	Power Problem		
2921	2014				CT257	Implantable cardiac pacemakers	E2401	Insufficient Information	A09	Output Problem		
									A0907	No Device Output		
2922	2014		47270	Cardiac resynchronization therapy implantable defibrillator	CT2866	Implantable cardiac electrical stimulators			A26	Insufficient Device Problem Information		
2923	2014		47270	Cardiac resynchronization therapy implantable defibrillator	CT2866	Implantable cardiac electrical stimulators			A26	Insufficient Device Problem Information		
2924	2014				CT257	Implantable cardiac pacemakers	E2401	Insufficient Information	A01	Patient Device Interaction Problem		
									A17	Compatibility Problem		
2925	2014	All other reportable incidents			CT257	Implantable cardiac pacemakers			A07	Electrical /Electronic Property Problem		
									A0708	Power Problem		
2926	2014		47270	Cardiac resynchronization therapy implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E0402	Hypersensitivity/Allergic Reaction	A26	Insufficient Device Problem Information		
2927	2014		35223	Endocardial/interventricular septal pacing lead	CT1684	Cardiac pacemakers and associated devices	F14	Prolonged Episode of Care	A15	Activation, Positioning or Separation Problem		
									A150101	Activation Failure		
2928	2014		47270	Cardiac resynchronization therapy implantable defibrillator	CT2866	Implantable cardiac electrical stimulators			A26	Insufficient Device Problem Information		
2929	2014		47270	Cardiac resynchronization therapy implantable defibrillator	CT2866	Implantable cardiac electrical stimulators			A26	Insufficient Device Problem Information		
2930	2014		47270	Cardiac resynchronization therapy implantable defibrillator	CT2866	Implantable cardiac electrical stimulators			A07	Electrical /Electronic Property Problem		
									A0709	Device Sensing Problem		
2931	2014				CT257	Implantable cardiac pacemakers	F1905	Device Revision or Replacement	A07	Electrical /Electronic Property Problem		
									A0708	Power Problem		
2932	2014				CT257	Implantable cardiac pacemakers	F1905	Device Revision or Replacement	A07	Electrical /Electronic Property Problem		
									A0709	Device Sensing Problem		
2933	2014				CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A07	Electrical /Electronic Property Problem		
									A0708	Power Problem		
2934	2014	All other reportable incidents			CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A07	Electrical /Electronic Property Problem		
									A0709	Device Sensing Problem		
2935	2014	All other reportable incidents			CT257	Implantable cardiac pacemakers	E2401	Insufficient Information	A09	Output Problem		
									A0907	No Device Output		
2936	2013	All other reportable incidents			CT257	Implantable cardiac pacemakers	F1905	Device Revision or Replacement	A07	Electrical /Electronic Property Problem		
									A0708	Power Problem		
2937	2013	All other reportable incidents			CT257	Implantable cardiac pacemakers	F1905	Device Revision or Replacement	A09	Output Problem	24/12/2012	
									A0907	No Device Output	24/12/2012	
									A1304	Telemetry Discrepancy	24/12/2012	
									A26	Insufficient Device Problem Information	24/12/2012	
2938	2013	All other reportable incidents			CT257	Implantable cardiac pacemakers	F1905	Device Revision or Replacement	A09	Output Problem	14/03/2013	
									A090404	Output Above Specifications	14/03/2013	
2939	2013		47270	Cardiac resynchronization therapy implantable defibrillator	CT2866	Implantable cardiac electrical stimulators			A26	Insufficient Device Problem Information		
2940	2013		47270	Cardiac resynchronization therapy implantable defibrillator	CT2866	Implantable cardiac electrical stimulators			A26	Insufficient Device Problem Information		
2941	2013		47270	Cardiac resynchronization therapy implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	F1905	Device Revision or Replacement	A07	Electrical /Electronic Property Problem		
									A0708	Power Problem		
2942	2013		47270	Cardiac resynchronization therapy implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	F15	Recognized Device or Procedural Complication	A26	Insufficient Device Problem Information		
2943	2014		47270	Cardiac resynchronization therapy implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A07	Electrical /Electronic Property Problem		
									A0709	Device Sensing Problem		
2944	2014	All other reportable incidents			CT257	Implantable cardiac pacemakers	F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A07	Electrical /Electronic Property Problem		
									A0709	Device Sensing Problem		
2945	2013		47270	Cardiac resynchronization therapy implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A13	Communication or Transmission Problem		
									A26	Insufficient Device Problem Information		
2946	2013		47270	Cardiac resynchronization therapy implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E2104	Electric Shock	A09	Output Problem		
									A090404	Output Above Specifications		
2947	2014				CT257	Implantable cardiac pacemakers	F14	Prolonged Episode of Care	A05	Mechanical Problem		
									A0506	Mechanical Jam		
									A17	Compatibility Problem		
									A1702	Device-Device Incompatibility		
2948	2014		47270	Cardiac resynchronization therapy implantable defibrillator	CT2866	Implantable cardiac electrical stimulators			A07	Electrical /Electronic Property Problem		
									A0709	Device Sensing Problem		
2949	2014		47270	Cardiac resynchronization therapy implantable defibrillator	CT2866	Implantable cardiac electrical stimulators			A07	Electrical /Electronic Property Problem		
									A0708	Power Problem		
2950	2014				CT257	Implantable cardiac pacemakers			A26	Insufficient Device Problem Information		
2951	2014	All other reportable incidents			CT257	Implantable cardiac pacemakers	F1905	Device Revision or Replacement	A07	Electrical /Electronic Property Problem	06/01/2014	
									A0708	Power Problem	06/01/2014	
2952	2014	All other reportable incidents			CT257	Implantable cardiac pacemakers	E2104	Electric Shock	A0104	Migration or Expulsion of Device		
							F14	Prolonged Episode of Care	A0104	Migration or Expulsion of Device		
2953	2013		61438	Cardiac pulse generator reader, home-use	CT1684	Cardiac pacemakers and associated devices	E0402	Hypersensitivity/Allergic Reaction	A10	Temperature Problem		
									A1009	Smoking		
2954	2014		47270	Cardiac resynchronization therapy implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	F21	Unexpected Deterioration	A26	Insufficient Device Problem Information		
2955	2014		47270	Cardiac resynchronization therapy implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	F21	Unexpected Deterioration	A26	Insufficient Device Problem Information		

2956	2014	Death	47270	Cardiac resynchronization therapy implantable defibrillator	CT2866	Implantable cardiac electrical stimulators			A09	Output Problem	04/02/2014	
									A0907	No Device Output	04/02/2014	
2957	2013	All other reportable incidents			CT257	Implantable cardiac pacemakers	F1905	Device Revision or Replacement	A07	Electrical /Electronic Property Problem	04/04/2013	
									A0708	Power Problem	04/04/2013	
2958	2014				CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A2001	Misassembled During Installation		
									A26	Insufficient Device Problem Information		
2959	2014	All other reportable incidents	47270	Cardiac resynchronization therapy implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A09	Output Problem		
									A0907	No Device Output		
2960	2014	All other reportable incidents			CT257	Implantable cardiac pacemakers	F1905	Device Revision or Replacement	A07	Electrical /Electronic Property Problem		
									A0708	Power Problem		
2961	2013		47270	Cardiac resynchronization therapy implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	F1905	Device Revision or Replacement	A07	Electrical /Electronic Property Problem		
									A0708	Power Problem		
2962	2014		47270	Cardiac resynchronization therapy implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E0402	Hypersensitivity/Allergic Reaction	A26	Insufficient Device Problem Information		
2963	2014		47270	Cardiac resynchronization therapy implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E2401	Insufficient Information	A26	Insufficient Device Problem Information		
2964	2014		47270	Cardiac resynchronization therapy implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E2401	Insufficient Information	A26	Insufficient Device Problem Information		
2965	2014		47270	Cardiac resynchronization therapy implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	F1905	Device Revision or Replacement	A07	Electrical /Electronic Property Problem		
									A0708	Power Problem		
2966	2014				CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A09	Output Problem		
									A0907	No Device Output		
2967	2013				CT257	Implantable cardiac pacemakers			A07	Electrical /Electronic Property Problem		
									A0708	Power Problem		
2968	2014				CT257	Implantable cardiac pacemakers	E0402	Hypersensitivity/Allergic Reaction	A26	Insufficient Device Problem Information		
2969	2014				CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A07	Electrical /Electronic Property Problem		
									A0708	Power Problem		
2970	2013		47270	Cardiac resynchronization therapy implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E2401	Insufficient Information	A26	Insufficient Device Problem Information		
2971	2013		47270	Cardiac resynchronization therapy implantable defibrillator	CT2866	Implantable cardiac electrical stimulators			A26	Insufficient Device Problem Information		
2972	2014	All other reportable incidents			CT257	Implantable cardiac pacemakers	E2401	Insufficient Information	A07	Electrical /Electronic Property Problem	05/02/2014	
									A0709	Device Sensing Problem	05/02/2014	
2973	2014	All other reportable incidents			CT257	Implantable cardiac pacemakers	F1905	Device Revision or Replacement	A07	Electrical /Electronic Property Problem	13/01/2014	
									A0708	Power Problem	13/01/2014	
2974	2014	All other reportable incidents			CT257	Implantable cardiac pacemakers	F1905	Device Revision or Replacement	A09	Output Problem		
									A0907	No Device Output		
									A13	Communication or Transmission Problem		
									A26	Insufficient Device Problem Information		
2975	2014	All other reportable incidents			CT257	Implantable cardiac pacemakers	F1905	Device Revision or Replacement	A09	Output Problem	31/01/2014	
									A0907	No Device Output	31/01/2014	
									A1304	Telemetry Discrepancy	31/01/2014	
									A26	Insufficient Device Problem Information	31/01/2014	
2976	2013		47270	Cardiac resynchronization therapy implantable defibrillator	CT2866	Implantable cardiac electrical stimulators			A26	Insufficient Device Problem Information		
2977	2014		47270	Cardiac resynchronization therapy implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E2401	Insufficient Information	A07	Electrical /Electronic Property Problem		
									A0709	Device Sensing Problem		
2978	2014		47270	Cardiac resynchronization therapy implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	F1905	Device Revision or Replacement	A07	Electrical /Electronic Property Problem		
									A0709	Device Sensing Problem		
2979	2014				CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A07	Electrical /Electronic Property Problem		
									A0709	Device Sensing Problem		
									A09	Output Problem		
									A0907	No Device Output		
2980	2013				CT1684	Cardiac pacemakers and associated devices	F21	Unexpected Deterioration	A040102	Loss of or Failure to Bond		
									A09	Output Problem		
									A0907	No Device Output		
									A26	Insufficient Device Problem Information		
2981	2013				CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A01	Patient Device Interaction Problem		
									A07	Electrical /Electronic Property Problem		
									A0709	Device Sensing Problem		
									A17	Compatibility Problem		
2982	2014	All other reportable incidents			CT257	Implantable cardiac pacemakers	F1905	Device Revision or Replacement	A07	Electrical /Electronic Property Problem	02/10/2014	
									A0708	Power Problem	02/10/2014	
2983	2014		35852	Single-chamber implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E1906	Unspecified Infection	A01	Patient Device Interaction Problem	26/10/2014	
									A0104	Migration or Expulsion of Device	26/10/2014	
									A17	Compatibility Problem	26/10/2014	
							F1905	Device Revision or Replacement	A01	Patient Device Interaction Problem	26/10/2014	
									A0104	Migration or Expulsion of Device	26/10/2014	
									A17	Compatibility Problem	26/10/2014	
2984	2013				CT257	Implantable cardiac pacemakers	F1905	Device Revision or Replacement	A09	Output Problem		
									A0907	No Device Output		
									A16	Protective Measures Problem		
									A1602	Fail-Safe Problem		
2985	2013				CT257	Implantable cardiac pacemakers	E2401	Insufficient Information	A11	Computer Software Problem		
									A1102	Application Program Problem		
2986	2013				CT257	Implantable cardiac pacemakers			A07	Electrical /Electronic Property Problem		
									A0709	Device Sensing Problem		
2987	2014		35852	Single-chamber implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E1906	Unspecified Infection	A01	Patient Device Interaction Problem		
									A17	Compatibility Problem		
							F1905	Device Revision or Replacement	A01	Patient Device Interaction Problem		
									A17	Compatibility Problem		
2988	2014	Death			CT257	Implantable cardiac pacemakers	E2104	Electric Shock	A07	Electrical /Electronic Property Problem		
									A0721	Circuit Failure		
2989	2014		47270	Cardiac resynchronization therapy implantable defibrillator	CT2866	Implantable cardiac electrical stimulators			A26	Insufficient Device Problem Information		
2990	2014		47270	Cardiac resynchronization therapy implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E0402	Hypersensitivity/Allergic Reaction	A26	Insufficient Device Problem Information		
2991	2014		47270	Cardiac resynchronization therapy implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	F21	Unexpected Deterioration	A26	Insufficient Device Problem Information		

2992	2014		47270	Cardiac resynchronization therapy implantable defibrillator	CT2866	Implantable cardiac electrical stimulators			A26	Insufficient Device Problem Information		
2993	2014				CT257	Implantable cardiac pacemakers			A07	Electrical /Electronic Property Problem		
									A0708	Power Problem		
2994	2014		47270	Cardiac resynchronization therapy implantable defibrillator	CT2866	Implantable cardiac electrical stimulators			A26	Insufficient Device Problem Information		
2995	2014	All other reportable incidents			CT257	Implantable cardiac pacemakers	F1905	Device Revision or Replacement	A09	Output Problem	17/11/2014	
									A0907	No Device Output	17/11/2014	
2996	2013		47270	Cardiac resynchronization therapy implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	F15	Recognized Device or Procedural Complication	A26	Insufficient Device Problem Information		
2997	2013		47270	Cardiac resynchronization therapy implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	F15	Recognized Device or Procedural Complication	A26	Insufficient Device Problem Information		
2998	2014				CT257	Implantable cardiac pacemakers	F1905	Device Revision or Replacement	A07	Electrical /Electronic Property Problem	05/11/2014	
									A0709	Device Sensing Problem	05/11/2014	
2999	2013				CT257	Implantable cardiac pacemakers	F15	Recognized Device or Procedural Complication	A26	Insufficient Device Problem Information		
3000	2013		47270	Cardiac resynchronization therapy implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	F15	Recognized Device or Procedural Complication	A26	Insufficient Device Problem Information		
3001	2013				CT257	Implantable cardiac pacemakers	F1905	Device Revision or Replacement	A07	Electrical /Electronic Property Problem		
									A0708	Power Problem		
3002	2014	All other reportable incidents			CT257	Implantable cardiac pacemakers	F1905	Device Revision or Replacement	A07	Electrical /Electronic Property Problem		
									A0708	Power Problem		
3003	2013	All other reportable incidents			CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A1501	Activation Problem		
									A23	Use of Device Problem		
3004	2014				CT257	Implantable cardiac pacemakers			A26	Insufficient Device Problem Information		
3005	2014		47270	Cardiac resynchronization therapy implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	F1905	Device Revision or Replacement	A07	Electrical /Electronic Property Problem		
									A0708	Power Problem		
3006	2014		35852	Single-chamber implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	F15	Recognized Device or Procedural Complication	A01	Patient Device Interaction Problem		
									A17	Compatibility Problem		
3007	2014	All other reportable incidents			CT257	Implantable cardiac pacemakers			A07	Electrical /Electronic Property Problem		
									A0708	Power Problem		
3008	2013				CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A09	Output Problem		
									A0908	Incorrect, Inadequate or Imprecise Result or Readings		
3009	2014		35852	Single-chamber implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E1906	Unspecified Infection	A01	Patient Device Interaction Problem	24/04/2014	
									A17	Compatibility Problem	24/04/2014	
3010	2013		47270	Cardiac resynchronization therapy implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A05	Mechanical Problem		
									A0504	Leak/Splash		
3011	2013		35223	Endocardial/interventricular septal pacing lead	CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A07	Electrical /Electronic Property Problem		
									A0709	Device Sensing Problem		
3012	2013		47270	Cardiac resynchronization therapy implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A11	Computer Software Problem		
									A1102	Application Program Problem		
3013	2014				CT257	Implantable cardiac pacemakers	F1905	Device Revision or Replacement	A07	Electrical /Electronic Property Problem		
									A0709	Device Sensing Problem		
									A09	Output Problem		
									A0907	No Device Output		
3014	2014				CT257	Implantable cardiac pacemakers			A07	Electrical /Electronic Property Problem		
									A0708	Power Problem		
3015	2014	All other reportable incidents			CT257	Implantable cardiac pacemakers	F1905	Device Revision or Replacement	A07	Electrical /Electronic Property Problem	10/04/2014	
									A0709	Device Sensing Problem	10/04/2014	
3016	2014	All other reportable incidents	47270	Cardiac resynchronization therapy implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	F1905	Device Revision or Replacement	A16	Protective Measures Problem	31/01/2014	
									A1602	Fail-Safe Problem	31/01/2014	
3017	2013		47270	Cardiac resynchronization therapy implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	F1905	Device Revision or Replacement	A07	Electrical /Electronic Property Problem	22/02/2013	
									A0708	Power Problem	22/02/2013	
3018	2014	All other reportable incidents	47270	Cardiac resynchronization therapy implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	F1905	Device Revision or Replacement	A07	Electrical /Electronic Property Problem	28/01/2014	
									A0708	Power Problem	28/01/2014	
3019	2013				CT257	Implantable cardiac pacemakers	F1905	Device Revision or Replacement	A07	Electrical /Electronic Property Problem		
									A0708	Power Problem		
3020	2014	All other reportable incidents			CT257	Implantable cardiac pacemakers	F1905	Device Revision or Replacement	A07	Electrical /Electronic Property Problem		
									A0708	Power Problem		
3021	2014	All other reportable incidents			CT257	Implantable cardiac pacemakers	F1905	Device Revision or Replacement	A07	Electrical /Electronic Property Problem		
									A0709	Device Sensing Problem		
3022	2013		47270	Cardiac resynchronization therapy implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	F21	Unexpected Deterioration	A26	Insufficient Device Problem Information		
3023	2014		47270	Cardiac resynchronization therapy implantable defibrillator	CT2866	Implantable cardiac electrical stimulators			A26	Insufficient Device Problem Information		
3024	2014		47270	Cardiac resynchronization therapy implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	F21	Unexpected Deterioration	A26	Insufficient Device Problem Information		
3025	2014		47270	Cardiac resynchronization therapy implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	F1905	Device Revision or Replacement	A07	Electrical /Electronic Property Problem		
									A0709	Device Sensing Problem		
3026	2014		47270	Cardiac resynchronization therapy implantable defibrillator	CT2866	Implantable cardiac electrical stimulators			A07	Electrical /Electronic Property Problem		
									A0708	Power Problem		
3027	2014		47270	Cardiac resynchronization therapy implantable defibrillator	CT2866	Implantable cardiac electrical stimulators			A26	Insufficient Device Problem Information		
3028	2014		47270	Cardiac resynchronization therapy implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E0402	Hypersensitivity/Allergic Reaction	A26	Insufficient Device Problem Information		
3029	2014	Death			CT257	Implantable cardiac pacemakers	E2401	Insufficient Information	A26	Insufficient Device Problem Information		
3030	2014				CT257	Implantable cardiac pacemakers	E0402	Hypersensitivity/Allergic Reaction	A26	Insufficient Device Problem Information		
3031	2013		47270	Cardiac resynchronization therapy implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A26	Insufficient Device Problem Information		
3032	2013	All other reportable incidents			CT257	Implantable cardiac pacemakers			A07	Electrical /Electronic Property Problem	22/04/2013	
									A0708	Power Problem	22/04/2013	
3033	2014				CT257	Implantable cardiac pacemakers	E0402	Hypersensitivity/Allergic Reaction	A26	Insufficient Device Problem Information		
3034	2014		47270	Cardiac resynchronization therapy implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E0402	Hypersensitivity/Allergic Reaction	A26	Insufficient Device Problem Information		
3035	2014		47270	Cardiac resynchronization therapy implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E0402	Hypersensitivity/Allergic Reaction	A26	Insufficient Device Problem Information		
3036	2014		47270	Cardiac resynchronization therapy implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E0402	Hypersensitivity/Allergic Reaction	A26	Insufficient Device Problem Information		
3037	2013				CT257	Implantable cardiac pacemakers	F1905	Device Revision or Replacement	A13	Communication or Transmission Problem		
									A26	Insufficient Device Problem Information		
3038	2013				CT257	Implantable cardiac pacemakers	F1905	Device Revision or Replacement	A07	Electrical /Electronic Property Problem		

									A0709	Device Sensing Problem		
3039	2014	All other reportable incidents	47270	Cardiac resynchronization therapy implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A1304	Telemetry Discrepancy		
									A26	Insufficient Device Problem Information		
3040	2013				CT257	Implantable cardiac pacemakers	F1905	Device Revision or Replacement	A09	Output Problem		
									A0907	No Device Output		
									A1304	Telemetry Discrepancy		
									A26	Insufficient Device Problem Information		
3041	2013				CT257	Implantable cardiac pacemakers			A26	Insufficient Device Problem Information		
3042	2013				CT257	Implantable cardiac pacemakers	F15	Recognized Device or Procedural Complication	A26	Insufficient Device Problem Information		
3043	2013		47270	Cardiac resynchronization therapy implantable defibrillator	CT2866	Implantable cardiac electrical stimulators			A26	Insufficient Device Problem Information		
3044	2014				CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A07	Electrical /Electronic Property Problem		
									A0708	Power Problem		
3045	2013		47270	Cardiac resynchronization therapy implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A09	Output Problem		
									A090404	Output Above Specifications		
3046	2014				CT257	Implantable cardiac pacemakers	F21	Unexpected Deterioration	A26	Insufficient Device Problem Information		
3047	2014		47270	Cardiac resynchronization therapy implantable defibrillator	CT2866	Implantable cardiac electrical stimulators			A26	Insufficient Device Problem Information		
3048	2013				CT257	Implantable cardiac pacemakers	E0752	Unspecified Respiratory Problem	A07	Electrical /Electronic Property Problem		
									A0708	Power Problem		
3049	2014		47270	Cardiac resynchronization therapy implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	F1905	Device Revision or Replacement	A07	Electrical /Electronic Property Problem	09/12/2014	
									A0708	Power Problem	09/12/2014	
3050	2013		47270	Cardiac resynchronization therapy implantable defibrillator	CT2866	Implantable cardiac electrical stimulators			A26	Insufficient Device Problem Information		
3051	2013		47270	Cardiac resynchronization therapy implantable defibrillator	CT2866	Implantable cardiac electrical stimulators			A26	Insufficient Device Problem Information		
3052	2014		47270	Cardiac resynchronization therapy implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	F21	Unexpected Deterioration	A26	Insufficient Device Problem Information		
3053	2014				CT257	Implantable cardiac pacemakers			A07	Electrical /Electronic Property Problem		
									A0709	Device Sensing Problem		
3054	2014				CT257	Implantable cardiac pacemakers	F1905	Device Revision or Replacement	A07	Electrical /Electronic Property Problem		
									A0708	Power Problem		
3055	2014	All other reportable incidents			CT257	Implantable cardiac pacemakers	F1905	Device Revision or Replacement	A07	Electrical /Electronic Property Problem	11/11/2014	
									A0708	Power Problem	11/11/2014	
3056	2014				CT1684	Cardiac pacemakers and associated devices			A07	Electrical /Electronic Property Problem		
									A0708	Power Problem		
3057	2014		47270	Cardiac resynchronization therapy implantable defibrillator	CT2866	Implantable cardiac electrical stimulators			A26	Insufficient Device Problem Information		
3058	2014		47270	Cardiac resynchronization therapy implantable defibrillator	CT2866	Implantable cardiac electrical stimulators			A26	Insufficient Device Problem Information		
3059	2014		35852	Single-chamber implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	F1905	Device Revision or Replacement	A01	Patient Device Interaction Problem		
									A17	Compatibility Problem		
3060	2014				CT257	Implantable cardiac pacemakers	E0402	Hypersensitivity/Allergic Reaction	A26	Insufficient Device Problem Information		
3061	2014		47270	Cardiac resynchronization therapy implantable defibrillator	CT2866	Implantable cardiac electrical stimulators			A26	Insufficient Device Problem Information		
3062	2014		47270	Cardiac resynchronization therapy implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A07	Electrical /Electronic Property Problem		
									A0708	Power Problem		
3063	2014				CT257	Implantable cardiac pacemakers	F21	Unexpected Deterioration	A26	Insufficient Device Problem Information		
3064	2014				CT257	Implantable cardiac pacemakers	F21	Unexpected Deterioration	A26	Insufficient Device Problem Information		
3065	2013	Unanticipated serious deterioration in state of health	35223	Endocardial/interventricular septal pacing lead	CT1684	Cardiac pacemakers and associated devices	E2114	Perforation	A01	Patient Device Interaction Problem		
									A17	Compatibility Problem		
3066	2013	All other reportable incidents			CT257	Implantable cardiac pacemakers	F1905	Device Revision or Replacement	A07	Electrical /Electronic Property Problem	30/07/2013	
									A0708	Power Problem	30/07/2013	
3067	2013	All other reportable incidents			CT257	Implantable cardiac pacemakers	F1905	Device Revision or Replacement	A07	Electrical /Electronic Property Problem		
									A0708	Power Problem		
3068	2014		47270	Cardiac resynchronization therapy implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E0402	Hypersensitivity/Allergic Reaction	A26	Insufficient Device Problem Information		
3069	2013	All other reportable incidents			CT257	Implantable cardiac pacemakers	F14	Prolonged Episode of Care	A07	Electrical /Electronic Property Problem		
									A0708	Power Problem		
							F1905	Device Revision or Replacement	A07	Electrical /Electronic Property Problem		
									A0708	Power Problem		
3070	2013		47270	Cardiac resynchronization therapy implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	F15	Recognized Device or Procedural Complication	A26	Insufficient Device Problem Information		
3071	2013		47270	Cardiac resynchronization therapy implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	F15	Recognized Device or Procedural Complication	A26	Insufficient Device Problem Information		
3072	2013		47270	Cardiac resynchronization therapy implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E0402	Hypersensitivity/Allergic Reaction	A26	Insufficient Device Problem Information		
3073	2014				CT257	Implantable cardiac pacemakers	F21	Unexpected Deterioration	A26	Insufficient Device Problem Information		
3074	2014	All other reportable incidents			CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A07	Electrical /Electronic Property Problem		
									A0708	Power Problem		
3075	2014	All other reportable incidents			CT257	Implantable cardiac pacemakers	F1905	Device Revision or Replacement	A07	Electrical /Electronic Property Problem	20/02/2014	
									A0708	Power Problem	20/02/2014	
3076	2014		47270	Cardiac resynchronization therapy implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E0402	Hypersensitivity/Allergic Reaction	A26	Insufficient Device Problem Information		
3077	2014				CT257	Implantable cardiac pacemakers	E0402	Hypersensitivity/Allergic Reaction	A26	Insufficient Device Problem Information		
3078	2014				CT257	Implantable cardiac pacemakers	F1905	Device Revision or Replacement	A07	Electrical /Electronic Property Problem		
									A0708	Power Problem		
3079	2014		47270	Cardiac resynchronization therapy implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E0402	Hypersensitivity/Allergic Reaction	A26	Insufficient Device Problem Information		
3080	2014		47270	Cardiac resynchronization therapy implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E0402	Hypersensitivity/Allergic Reaction	A26	Insufficient Device Problem Information		
3081	2014				CT257	Implantable cardiac pacemakers	F1905	Device Revision or Replacement	A07	Electrical /Electronic Property Problem		
									A0708	Power Problem		
3082	2013				CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A07	Electrical /Electronic Property Problem		
									A0709	Device Sensing Problem		
									A13	Communication or Transmission Problem		
									A26	Insufficient Device Problem Information		
3083	2013		47270	Cardiac resynchronization therapy implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	F15	Recognized Device or Procedural Complication	A26	Insufficient Device Problem Information		
3084	2013		47270	Cardiac resynchronization therapy implantable defibrillator	CT2866	Implantable cardiac electrical stimulators			A26	Insufficient Device Problem Information		

3085	2013		47270	Cardiac resynchronization therapy implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	F21	Unexpected Deterioration	A26	Insufficient Device Problem Information		
3086	2014	All other reportable incidents			CT257	Implantable cardiac pacemakers	F1905	Device Revision or Replacement	A11	Computer Software Problem		
									A1102	Application Program Problem		
3087	2014				CT257	Implantable cardiac pacemakers	F1905	Device Revision or Replacement	A07	Electrical /Electronic Property Problem		
									A0708	Power Problem		
3088	2014	All other reportable incidents			CT257	Implantable cardiac pacemakers	F1905	Device Revision or Replacement	A07	Electrical /Electronic Property Problem		
									A0709	Device Sensing Problem		
3089	2014	All other reportable incidents			CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A07	Electrical /Electronic Property Problem		
									A0708	Power Problem		
3090	2013	All other reportable incidents			CT257	Implantable cardiac pacemakers	F1905	Device Revision or Replacement	A07	Electrical /Electronic Property Problem	02/08/2013	
									A0708	Power Problem	02/08/2013	
3091	2014				CT257	Implantable cardiac pacemakers			A26	Insufficient Device Problem Information		
3092	2014		47270	Cardiac resynchronization therapy implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E0402	Hypersensitivity/Allergic Reaction	A26	Insufficient Device Problem Information		
3093	2013		47270	Cardiac resynchronization therapy implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	F1905	Device Revision or Replacement	A07	Electrical /Electronic Property Problem		
									A0709	Device Sensing Problem		
3094	2013				CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A07	Electrical /Electronic Property Problem		
									A0708	Power Problem		
3095	2013				CT257	Implantable cardiac pacemakers	F1905	Device Revision or Replacement	A07	Electrical /Electronic Property Problem		
									A0708	Power Problem		
3096	2014				CT257	Implantable cardiac pacemakers	F1905	Device Revision or Replacement	A07	Electrical /Electronic Property Problem		
									A0709	Device Sensing Problem		
3097	2014		35223	Endocardial/interventricular septal pacing lead	CT1684	Cardiac pacemakers and associated devices	F15	Recognized Device or Procedural Complication	A04	Material Integrity Problem		
									A0413	Material Separation		
3098	2013				CT257	Implantable cardiac pacemakers	F1905	Device Revision or Replacement	A07	Electrical /Electronic Property Problem		
									A0708	Power Problem		
3099	2013		47270	Cardiac resynchronization therapy implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	F1905	Device Revision or Replacement	A07	Electrical /Electronic Property Problem		
									A0708	Power Problem		
3100	2013				CT257	Implantable cardiac pacemakers	F1905	Device Revision or Replacement	A07	Electrical /Electronic Property Problem		
									A0708	Power Problem		
3101	2013				CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A07	Electrical /Electronic Property Problem		
									A0703	Arcing		
									A0721	Circuit Failure		
3102	2014	All other reportable incidents	47270	Cardiac resynchronization therapy implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	F1905	Device Revision or Replacement	A07	Electrical /Electronic Property Problem	04/03/2014	
									A0709	Device Sensing Problem	04/03/2014	
3103	2013		35223	Endocardial/interventricular septal pacing lead	CT1684	Cardiac pacemakers and associated devices	E2104	Electric Shock	A07	Electrical /Electronic Property Problem		
									A0709	Device Sensing Problem		
3104	2014		47270	Cardiac resynchronization therapy implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	F21	Unexpected Deterioration	A26	Insufficient Device Problem Information		
3105	2014		47270	Cardiac resynchronization therapy implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E0402	Hypersensitivity/Allergic Reaction	A26	Insufficient Device Problem Information		
3106	2014		47270	Cardiac resynchronization therapy implantable defibrillator	CT2866	Implantable cardiac electrical stimulators			A26	Insufficient Device Problem Information		
3107	2014		47270	Cardiac resynchronization therapy implantable defibrillator	CT2866	Implantable cardiac electrical stimulators			A26	Insufficient Device Problem Information		
3108	2013		47270	Cardiac resynchronization therapy implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E0402	Hypersensitivity/Allergic Reaction	A26	Insufficient Device Problem Information		
3109	2013		47270	Cardiac resynchronization therapy implantable defibrillator	CT2866	Implantable cardiac electrical stimulators			A26	Insufficient Device Problem Information		
3110	2013		47270	Cardiac resynchronization therapy implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	F21	Unexpected Deterioration	A26	Insufficient Device Problem Information		
3111	2014				CT257	Implantable cardiac pacemakers	F15	Recognized Device or Procedural Complication	A09	Output Problem		
									A0907	No Device Output		
3112	2013				CT257	Implantable cardiac pacemakers	F1905	Device Revision or Replacement	A1304	Telemetry Discrepancy		
									A26	Insufficient Device Problem Information		
3113	2014	All other reportable incidents	47270	Cardiac resynchronization therapy implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A16	Protective Measures Problem		
									A1602	Fail-Safe Problem		
3114	2013	All other reportable incidents			CT257	Implantable cardiac pacemakers	F1905	Device Revision or Replacement	A07	Electrical /Electronic Property Problem	31/07/2013	
									A0708	Power Problem	31/07/2013	
3115	2014				CT257	Implantable cardiac pacemakers			A26	Insufficient Device Problem Information		
3116	2014				CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A26	Insufficient Device Problem Information		
3117	2014		47270	Cardiac resynchronization therapy implantable defibrillator	CT2866	Implantable cardiac electrical stimulators			A26	Insufficient Device Problem Information		
3118	2014		47270	Cardiac resynchronization therapy implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E2401	Insufficient Information	A26	Insufficient Device Problem Information		
3119	2013				CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A13	Communication or Transmission Problem		
									A26	Insufficient Device Problem Information		
3120	2014	All other reportable incidents			CT257	Implantable cardiac pacemakers	F1905	Device Revision or Replacement	A07	Electrical /Electronic Property Problem	24/02/2014	
									A0709	Device Sensing Problem	24/02/2014	
									A09	Output Problem	24/02/2014	
									A0907	No Device Output	24/02/2014	
3121	2014	All other reportable incidents			CT257	Implantable cardiac pacemakers			A07	Electrical /Electronic Property Problem		
									A0709	Device Sensing Problem		
3122	2013		47270	Cardiac resynchronization therapy implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	F15	Recognized Device or Procedural Complication	A26	Insufficient Device Problem Information		
3123	2013		47270	Cardiac resynchronization therapy implantable defibrillator	CT2866	Implantable cardiac electrical stimulators			A26	Insufficient Device Problem Information		
3124	2013		47270	Cardiac resynchronization therapy implantable defibrillator	CT2866	Implantable cardiac electrical stimulators			A26	Insufficient Device Problem Information		
3125	2013		47270	Cardiac resynchronization therapy implantable defibrillator	CT2866	Implantable cardiac electrical stimulators			A26	Insufficient Device Problem Information		
3126	2013	All other reportable incidents	47270	Cardiac resynchronization therapy implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	F1905	Device Revision or Replacement	A07	Electrical /Electronic Property Problem	16/08/2013	
									A0708	Power Problem	16/08/2013	
3127	2013	All other reportable incidents			CT257	Implantable cardiac pacemakers			A09	Output Problem		
									A0907	No Device Output		
									A1304	Telemetry Discrepancy		
									A26	Insufficient Device Problem Information		
3128	2014		47270	Cardiac resynchronization therapy implantable defibrillator	CT2866	Implantable cardiac electrical stimulators			A26	Insufficient Device Problem Information		
3129	2014				CT257	Implantable cardiac pacemakers	E0402	Hypersensitivity/Allergic Reaction	A26	Insufficient Device Problem Information		
3130	2014				CT257	Implantable cardiac pacemakers	F1905	Device Revision or Replacement	A07	Electrical /Electronic Property Problem		

									A0709	Device Sensing Problem		
3131	2014		47270	Cardiac resynchronization therapy implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	F21	Unexpected Deterioration	A26	Insufficient Device Problem Information		
3132	2013				CT257	Implantable cardiac pacemakers	E2401	Insufficient Information	A07	Electrical /Electronic Property Problem		
									A0708	Power Problem		
3133	2014				CT257	Implantable cardiac pacemakers	F1905	Device Revision or Replacement	A07	Electrical /Electronic Property Problem		
									A0709	Device Sensing Problem		
3134	2014	All other reportable incidents			CT257	Implantable cardiac pacemakers	E2401	Insufficient Information	A07	Electrical /Electronic Property Problem		
									A0709	Device Sensing Problem		
3135	2014				CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A09	Output Problem		
									A0907	No Device Output		
3136	2014		47270	Cardiac resynchronization therapy implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E0402	Hypersensitivity/Allergic Reaction	A26	Insufficient Device Problem Information		
3137	2014		47270	Cardiac resynchronization therapy implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E0402	Hypersensitivity/Allergic Reaction	A26	Insufficient Device Problem Information		
3138	2013	All other reportable incidents	35223	Endocardial/interventricular septal pacing lead	CT1684	Cardiac pacemakers and associated devices	E0139	Unspecified Nervous System Problem	A07	Electrical /Electronic Property Problem		
									A0709	Device Sensing Problem		
3139	2013		47270	Cardiac resynchronization therapy implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	F1905	Device Revision or Replacement	A07	Electrical /Electronic Property Problem		
									A0709	Device Sensing Problem		
3140	2013		47270	Cardiac resynchronization therapy implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	F1905	Device Revision or Replacement	A07	Electrical /Electronic Property Problem		
									A0708	Power Problem		
3141	2014	All other reportable incidents	35223	Endocardial/interventricular septal pacing lead	CT1684	Cardiac pacemakers and associated devices	F05	Delay to Treatment/ Therapy	A07	Electrical /Electronic Property Problem		
									A0709	Device Sensing Problem		
3142	2014	All other reportable incidents			CT257	Implantable cardiac pacemakers	F1905	Device Revision or Replacement	A07	Electrical /Electronic Property Problem	03/06/2014	
									A0708	Power Problem	03/06/2014	
3143	2014	All other reportable incidents			CT257	Implantable cardiac pacemakers	F1905	Device Revision or Replacement	A07	Electrical /Electronic Property Problem		
									A0708	Power Problem		
3144	2014				CT257	Implantable cardiac pacemakers	F1905	Device Revision or Replacement	A07	Electrical /Electronic Property Problem	05/12/2014	
									A0708	Power Problem	05/12/2014	
3145	2014				CT257	Implantable cardiac pacemakers	F1905	Device Revision or Replacement	A07	Electrical /Electronic Property Problem		
									A0709	Device Sensing Problem		
3146	2013		47270	Cardiac resynchronization therapy implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	F15	Recognized Device or Procedural Complication	A26	Insufficient Device Problem Information		
3147	2014		47270	Cardiac resynchronization therapy implantable defibrillator	CT2866	Implantable cardiac electrical stimulators			A26	Insufficient Device Problem Information		
3148	2014		47270	Cardiac resynchronization therapy implantable defibrillator	CT2866	Implantable cardiac electrical stimulators			A26	Insufficient Device Problem Information		
3149	2014		47270	Cardiac resynchronization therapy implantable defibrillator	CT2866	Implantable cardiac electrical stimulators			A26	Insufficient Device Problem Information		
3150	2013		47270	Cardiac resynchronization therapy implantable defibrillator	CT2866	Implantable cardiac electrical stimulators			A26	Insufficient Device Problem Information		
3151	2013		47270	Cardiac resynchronization therapy implantable defibrillator	CT2866	Implantable cardiac electrical stimulators			A26	Insufficient Device Problem Information		
3152	2014		47270	Cardiac resynchronization therapy implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E0402	Hypersensitivity/Allergic Reaction	A26	Insufficient Device Problem Information		
3153	2014				CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A13	Communication or Transmission Problem		
									A26	Insufficient Device Problem Information		
3154	2014				CT257	Implantable cardiac pacemakers	E0402	Hypersensitivity/Allergic Reaction	A26	Insufficient Device Problem Information		
3155	2013	All other reportable incidents			CT257	Implantable cardiac pacemakers			A07	Electrical /Electronic Property Problem	27/08/2013	
									A0708	Power Problem	27/08/2013	
3156	2014	All other reportable incidents			CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A07	Electrical /Electronic Property Problem		
									A0721	Circuit Failure		
3157	2014		47270	Cardiac resynchronization therapy implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E0402	Hypersensitivity/Allergic Reaction	A26	Insufficient Device Problem Information		
3158	2014		47270	Cardiac resynchronization therapy implantable defibrillator	CT2866	Implantable cardiac electrical stimulators			A0104	Migration or Expulsion of Device		
									A07	Electrical /Electronic Property Problem		
									A0709	Device Sensing Problem		
3159	2014		47270	Cardiac resynchronization therapy implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E1906	Unspecified Infection	A01	Patient Device Interaction Problem		
									A17	Compatibility Problem		
							E2109	Implant Pain	A01	Patient Device Interaction Problem		
									A17	Compatibility Problem		
3160	2013	All other reportable incidents	47270	Cardiac resynchronization therapy implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	F1905	Device Revision or Replacement	A07	Electrical /Electronic Property Problem	17/02/2013	
									A0708	Power Problem	17/02/2013	
3161	2013		47270	Cardiac resynchronization therapy implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	F1905	Device Revision or Replacement	A07	Electrical /Electronic Property Problem		
									A0709	Device Sensing Problem		
3162	2013	All other reportable incidents			CT257	Implantable cardiac pacemakers	F1905	Device Revision or Replacement	A07	Electrical /Electronic Property Problem	13/05/2013	
									A0708	Power Problem	13/05/2013	
3163	2014	All other reportable incidents	35223	Endocardial/interventricular septal pacing lead	CT1684	Cardiac pacemakers and associated devices	F1905	Device Revision or Replacement	A07	Electrical /Electronic Property Problem	01/10/2010	
									A0709	Device Sensing Problem	01/10/2010	
									A1304	Telemetry Discrepancy	01/10/2010	
									A26	Insufficient Device Problem Information	01/10/2010	
3164	2014		47270	Cardiac resynchronization therapy implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E0402	Hypersensitivity/Allergic Reaction	A26	Insufficient Device Problem Information		
3165	2013				CT257	Implantable cardiac pacemakers	F1905	Device Revision or Replacement	A07	Electrical /Electronic Property Problem		
									A0708	Power Problem		
3166	2014				CT257	Implantable cardiac pacemakers	F1905	Device Revision or Replacement	A07	Electrical /Electronic Property Problem		
									A0708	Power Problem		
3167	2013		47270	Cardiac resynchronization therapy implantable defibrillator	CT2866	Implantable cardiac electrical stimulators			A26	Insufficient Device Problem Information		
3168	2013		47270	Cardiac resynchronization therapy implantable defibrillator	CT2866	Implantable cardiac electrical stimulators			A26	Insufficient Device Problem Information		
3169	2013		47270	Cardiac resynchronization therapy implantable defibrillator	CT2866	Implantable cardiac electrical stimulators			A26	Insufficient Device Problem Information		
3170	2013		47270	Cardiac resynchronization therapy implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	F15	Recognized Device or Procedural Complication	A26	Insufficient Device Problem Information		
3171	2014				CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A07	Electrical /Electronic Property Problem		
									A0708	Power Problem		
3172	2013				CT257	Implantable cardiac pacemakers			A26	Insufficient Device Problem Information		
3173	2014		47270	Cardiac resynchronization therapy implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E0402	Hypersensitivity/Allergic Reaction	A26	Insufficient Device Problem Information		
3174	2013	All other reportable incidents	35223	Endocardial/interventricular septal pacing lead	CT1684	Cardiac pacemakers and associated devices	E2401	Insufficient Information	A07	Electrical /Electronic Property Problem		
									A0709	Device Sensing Problem		

									A09	Output Problem		
									A0907	No Device Output		
3175	2014		47270	Cardiac resynchronization therapy implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A07	Electrical /Electronic Property Problem		
									A0709	Device Sensing Problem		
3176	2014				CT257	Implantable cardiac pacemakers	E0402	Hypersensitivity/Allergic Reaction	A26	Insufficient Device Problem Information		
3177	2013		47270	Cardiac resynchronization therapy implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E2401	Insufficient Information	A07	Electrical /Electronic Property Problem		
									A0709	Device Sensing Problem		
3178	2014				CT257	Implantable cardiac pacemakers	F1905	Device Revision or Replacement	A07	Electrical /Electronic Property Problem		
									A0708	Power Problem		
3179	2014		47270	Cardiac resynchronization therapy implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	F14	Prolonged Episode of Care	A13	Communication or Transmission Problem		
									A26	Insufficient Device Problem Information		
							F1905	Device Revision or Replacement	A13	Communication or Transmission Problem		
									A26	Insufficient Device Problem Information		
3180	2013				CT257	Implantable cardiac pacemakers	E2104	Electric Shock	A09	Output Problem		
									A090404	Output Above Specifications		
									A1304	Telemetry Discrepancy		
									A26	Insufficient Device Problem Information		
3181	2013				CT257	Implantable cardiac pacemakers	F1905	Device Revision or Replacement	A13	Communication or Transmission Problem		
									A16	Protective Measures Problem		
									A1602	Fail-Safe Problem		
									A26	Insufficient Device Problem Information		
3182	2014				CT257	Implantable cardiac pacemakers			A26	Insufficient Device Problem Information		
3183	2014				CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A07	Electrical /Electronic Property Problem		
									A0708	Power Problem		
3184	2014				CT257	Implantable cardiac pacemakers	E2401	Insufficient Information	A07	Electrical /Electronic Property Problem		
									A0708	Power Problem		
3185	2013		47270	Cardiac resynchronization therapy implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	F15	Recognized Device or Procedural Complication	A26	Insufficient Device Problem Information		
3186	2014				CT257	Implantable cardiac pacemakers	F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A16	Protective Measures Problem		
									A1602	Fail-Safe Problem		
3187	2013		47270	Cardiac resynchronization therapy implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A1304	Telemetry Discrepancy		
									A26	Insufficient Device Problem Information		
3188	2013	All other reportable incidents			CT257	Implantable cardiac pacemakers	E2401	Insufficient Information	A13	Communication or Transmission Problem		
									A26	Insufficient Device Problem Information		
3189	2013	All other reportable incidents			CT257	Implantable cardiac pacemakers	F1905	Device Revision or Replacement	A07	Electrical /Electronic Property Problem	15/05/2013	
									A0708	Power Problem	15/05/2013	
3190	2013	All other reportable incidents			CT257	Implantable cardiac pacemakers			A07	Electrical /Electronic Property Problem		
									A0708	Power Problem		
3191	2014				CT257	Implantable cardiac pacemakers	E0139	Unspecified Nervous System Problem	A07	Electrical /Electronic Property Problem		
									A0708	Power Problem		
									A09	Output Problem		
									A090404	Output Above Specifications		
							F1905	Device Revision or Replacement	A07	Electrical /Electronic Property Problem		
									A0708	Power Problem		
									A09	Output Problem		
									A090404	Output Above Specifications		
3192	2014		47270	Cardiac resynchronization therapy implantable defibrillator	CT2866	Implantable cardiac electrical stimulators			A26	Insufficient Device Problem Information		
3193	2013		47270	Cardiac resynchronization therapy implantable defibrillator	CT2866	Implantable cardiac electrical stimulators			A26	Insufficient Device Problem Information		
3194	2014		35223	Endocardial/interventricular septal pacing lead	CT1684	Cardiac pacemakers and associated devices	F1905	Device Revision or Replacement	A07	Electrical /Electronic Property Problem		
									A0709	Device Sensing Problem		
3195	2013		47270	Cardiac resynchronization therapy implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A07	Electrical /Electronic Property Problem		
									A0708	Power Problem		
3196	2013		47270	Cardiac resynchronization therapy implantable defibrillator	CT2866	Implantable cardiac electrical stimulators			A26	Insufficient Device Problem Information		
3197	2013		47270	Cardiac resynchronization therapy implantable defibrillator	CT2866	Implantable cardiac electrical stimulators			A26	Insufficient Device Problem Information		
3198	2013		47270	Cardiac resynchronization therapy implantable defibrillator	CT2866	Implantable cardiac electrical stimulators			A26	Insufficient Device Problem Information		
3199	2013		47270	Cardiac resynchronization therapy implantable defibrillator	CT2866	Implantable cardiac electrical stimulators			A26	Insufficient Device Problem Information		
3200	2013		47270	Cardiac resynchronization therapy implantable defibrillator	CT2866	Implantable cardiac electrical stimulators			A26	Insufficient Device Problem Information		
3201	2013		35223	Endocardial/interventricular septal pacing lead	CT1684	Cardiac pacemakers and associated devices	E2401	Insufficient Information	A04	Material Integrity Problem		
									A0404	Crack		
3202	2014		47270	Cardiac resynchronization therapy implantable defibrillator	CT2866	Implantable cardiac electrical stimulators			A26	Insufficient Device Problem Information		
3203	2014		47270	Cardiac resynchronization therapy implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E2401	Insufficient Information	A26	Insufficient Device Problem Information		
3204	2014		47270	Cardiac resynchronization therapy implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E0402	Hypersensitivity/Allergic Reaction	A26	Insufficient Device Problem Information		
3205	2013	All other reportable incidents	35223	Endocardial/interventricular septal pacing lead	CT1684	Cardiac pacemakers and associated devices	F1905	Device Revision or Replacement	A07	Electrical /Electronic Property Problem		
									A0709	Device Sensing Problem		
3206	2013		47270	Cardiac resynchronization therapy implantable defibrillator	CT2866	Implantable cardiac electrical stimulators			A26	Insufficient Device Problem Information		
3207	2013				CT257	Implantable cardiac pacemakers	F1905	Device Revision or Replacement	A07	Electrical /Electronic Property Problem		
									A0708	Power Problem		
							F21	Unexpected Deterioration	A07	Electrical /Electronic Property Problem		
									A0708	Power Problem		
3208	2013		47270	Cardiac resynchronization therapy implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	F1905	Device Revision or Replacement	A07	Electrical /Electronic Property Problem		
									A0708	Power Problem		
3209	2013		47270	Cardiac resynchronization therapy implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A09	Output Problem		
									A0908	Incorrect, Inadequate or Imprecise Result or Readings		
3210	2013	All other reportable incidents			CT257	Implantable cardiac pacemakers	F1905	Device Revision or Replacement	A07	Electrical /Electronic Property Problem	07/06/2013	
									A0708	Power Problem	07/06/2013	
3211	2013	All other reportable incidents			CT257	Implantable cardiac pacemakers	F1905	Device Revision or Replacement	A07	Electrical /Electronic Property Problem		
									A0708	Power Problem		
3212	2013		47270	Cardiac resynchronization therapy implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	F1905	Device Revision or Replacement	A07	Electrical /Electronic Property Problem		
									A0708	Power Problem		

3213	2014	All other reportable incidents			CT257	Implantable cardiac pacemakers	F1905	Device Revision or Replacement	A07	Electrical /Electronic Property Problem	16/06/2014	
									A0708	Power Problem	16/06/2014	
3214	2014	All other reportable incidents			CT257	Implantable cardiac pacemakers			A07	Electrical /Electronic Property Problem		
									A0709	Device Sensing Problem		
3215	2013				CT257	Implantable cardiac pacemakers	F1905	Device Revision or Replacement	A07	Electrical /Electronic Property Problem		
									A0709	Device Sensing Problem		
3216	2014		47270	Cardiac resynchronization therapy implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	F1905	Device Revision or Replacement	A07	Electrical /Electronic Property Problem		
									A0708	Power Problem		
									A13	Communication or Transmission Problem		
									A26	Insufficient Device Problem Information		
3217	2014				CT257	Implantable cardiac pacemakers	E2401	Insufficient Information	A26	Insufficient Device Problem Information		
3218	2013				CT257	Implantable cardiac pacemakers	F1905	Device Revision or Replacement	A07	Electrical /Electronic Property Problem		
									A0708	Power Problem		
3219	2013				CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A11	Computer Software Problem		
									A1102	Application Program Problem		
3220	2013		47270	Cardiac resynchronization therapy implantable defibrillator	CT2866	Implantable cardiac electrical stimulators			A07	Electrical /Electronic Property Problem		
									A0709	Device Sensing Problem		
3221	2013				CT257	Implantable cardiac pacemakers	F1905	Device Revision or Replacement	A07	Electrical /Electronic Property Problem		
									A0721	Circuit Failure		
3222	2014	All other reportable incidents	35223	Endocardial/interventricular septal pacing lead	CT1684	Cardiac pacemakers and associated devices	F1905	Device Revision or Replacement	A07	Electrical /Electronic Property Problem		
									A0709	Device Sensing Problem		
3223	2014	All other reportable incidents			CT257	Implantable cardiac pacemakers			A07	Electrical /Electronic Property Problem		
									A0708	Power Problem		
3224	2014	All other reportable incidents	35223	Endocardial/interventricular septal pacing lead	CT1684	Cardiac pacemakers and associated devices			A07	Electrical /Electronic Property Problem		
									A0709	Device Sensing Problem		
3225	2013		47270	Cardiac resynchronization therapy implantable defibrillator	CT2866	Implantable cardiac electrical stimulators			A26	Insufficient Device Problem Information		
3226	2014				CT257	Implantable cardiac pacemakers			A07	Electrical /Electronic Property Problem		
									A0708	Power Problem		
3227	2013		47270	Cardiac resynchronization therapy implantable defibrillator	CT2866	Implantable cardiac electrical stimulators			A26	Insufficient Device Problem Information		
3228	2013		47270	Cardiac resynchronization therapy implantable defibrillator	CT2866	Implantable cardiac electrical stimulators			A26	Insufficient Device Problem Information		
3229	2013		47270	Cardiac resynchronization therapy implantable defibrillator	CT2866	Implantable cardiac electrical stimulators			A26	Insufficient Device Problem Information		
3230	2013		47270	Cardiac resynchronization therapy implantable defibrillator	CT2866	Implantable cardiac electrical stimulators			A26	Insufficient Device Problem Information		
3231	2014		47270	Cardiac resynchronization therapy implantable defibrillator	CT2866	Implantable cardiac electrical stimulators			A26	Insufficient Device Problem Information		
3232	2014		47270	Cardiac resynchronization therapy implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E2401	Insufficient Information	A26	Insufficient Device Problem Information		
3233	2014	All other reportable incidents			CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A15	Activation, Positioning or Separation Problem		
									A150101	Activation Failure		
3234	2014				CT257	Implantable cardiac pacemakers	E2401	Insufficient Information	A26	Insufficient Device Problem Information		
3235	2014				CT257	Implantable cardiac pacemakers	E0402	Hypersensitivity/Allergic Reaction	A26	Insufficient Device Problem Information		
3236	2013		47270	Cardiac resynchronization therapy implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	F1905	Device Revision or Replacement	A07	Electrical /Electronic Property Problem		
									A0708	Power Problem		
3237	2013		47270	Cardiac resynchronization therapy implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	F1905	Device Revision or Replacement	A07	Electrical /Electronic Property Problem		
									A0708	Power Problem		
3238	2014	All other reportable incidents			CT257	Implantable cardiac pacemakers	E2401	Insufficient Information	A07	Electrical /Electronic Property Problem		
									A0708	Power Problem		
3239	2014	All other reportable incidents			CT257	Implantable cardiac pacemakers	F1905	Device Revision or Replacement	A07	Electrical /Electronic Property Problem	18/06/2014	
									A0708	Power Problem	18/06/2014	
3240	2014	All other reportable incidents			CT257	Implantable cardiac pacemakers	F1905	Device Revision or Replacement	A11	Computer Software Problem		
									A1102	Application Program Problem		
3241	2014	All other reportable incidents	47270	Cardiac resynchronization therapy implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	F1905	Device Revision or Replacement	A07	Electrical /Electronic Property Problem	09/07/2014	
									A0708	Power Problem	09/07/2014	
3242	2013				CT257	Implantable cardiac pacemakers			A11	Computer Software Problem		
									A1102	Application Program Problem		
3243	2014				CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A07	Electrical /Electronic Property Problem		
									A0708	Power Problem		
3244	2013	All other reportable incidents	35223	Endocardial/interventricular septal pacing lead	CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A04	Material Integrity Problem		
									A0405	Degraded		
3245	2013		47270	Cardiac resynchronization therapy implantable defibrillator	CT2866	Implantable cardiac electrical stimulators			A26	Insufficient Device Problem Information		
3246	2013		47270	Cardiac resynchronization therapy implantable defibrillator	CT2866	Implantable cardiac electrical stimulators			A26	Insufficient Device Problem Information		
3247	2014		47270	Cardiac resynchronization therapy implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E0402	Hypersensitivity/Allergic Reaction	A26	Insufficient Device Problem Information		
3248	2013		47270	Cardiac resynchronization therapy implantable defibrillator	CT2866	Implantable cardiac electrical stimulators			A26	Insufficient Device Problem Information		
3249	2014		47270	Cardiac resynchronization therapy implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E0402	Hypersensitivity/Allergic Reaction	A26	Insufficient Device Problem Information		
3250	2014	All other reportable incidents			CT257	Implantable cardiac pacemakers	E2401	Insufficient Information	A13	Communication or Transmission Problem	30/06/2014	
									A1304	Telemetry Discrepancy	30/06/2014	
									A26	Insufficient Device Problem Information	30/06/2014	
3251	2014		47270	Cardiac resynchronization therapy implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E0402	Hypersensitivity/Allergic Reaction	A26	Insufficient Device Problem Information		
3252	2014		47270	Cardiac resynchronization therapy implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E0402	Hypersensitivity/Allergic Reaction	A26	Insufficient Device Problem Information		
3253	2014		47270	Cardiac resynchronization therapy implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E0402	Hypersensitivity/Allergic Reaction	A26	Insufficient Device Problem Information		
3254	2014		47270	Cardiac resynchronization therapy implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E0402	Hypersensitivity/Allergic Reaction	A26	Insufficient Device Problem Information		
3255	2014		35223	Endocardial/interventricular septal pacing lead	CT1684	Cardiac pacemakers and associated devices			A26	Insufficient Device Problem Information		
3256	2014	All other reportable incidents			CT257	Implantable cardiac pacemakers			A13	Communication or Transmission Problem	26/06/2014	
									A26	Insufficient Device Problem Information	26/06/2014	
3257	2013				CT257	Implantable cardiac pacemakers	F1905	Device Revision or Replacement	A07	Electrical /Electronic Property Problem	01/10/2013	
									A0708	Power Problem	01/10/2013	
3258	2013	All other reportable incidents			CT257	Implantable cardiac pacemakers	F1905	Device Revision or Replacement	A13	Communication or Transmission Problem		
									A16	Protective Measures Problem		

									A1602	Fail-Safe Problem		
									A26	Insufficient Device Problem Information		
3259	2013		47263	Cardiac resynchronization therapy implantable pacemaker	CT1684	Cardiac pacemakers and associated devices	F15	Recognized Device or Procedural Complication	A26	Insufficient Device Problem Information		
					CT1685	Cardiac pacemakers	F15	Recognized Device or Procedural Complication	A26	Insufficient Device Problem Information		
					CT257	Implantable cardiac pacemakers	F15	Recognized Device or Procedural Complication	A26	Insufficient Device Problem Information		
					CT2866	Implantable cardiac electrical stimulators	F15	Recognized Device or Procedural Complication	A26	Insufficient Device Problem Information		
3260	2013		47270	Cardiac resynchronization therapy implantable defibrillator	CT2866	Implantable cardiac electrical stimulators			A26	Insufficient Device Problem Information		
3261	2013		47270	Cardiac resynchronization therapy implantable defibrillator	CT2866	Implantable cardiac electrical stimulators			A26	Insufficient Device Problem Information		
3262	2014	All other reportable incidents			CT257	Implantable cardiac pacemakers			A07	Electrical /Electronic Property Problem		
									A0708	Power Problem		
3263	2013		47270	Cardiac resynchronization therapy implantable defibrillator	CT2866	Implantable cardiac electrical stimulators			A26	Insufficient Device Problem Information		
3264	2014		47270	Cardiac resynchronization therapy implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E0402	Hypersensitivity/Allergic Reaction	A26	Insufficient Device Problem Information		
3265	2014		47263	Cardiac resynchronization therapy implantable pacemaker	CT1684	Cardiac pacemakers and associated devices	F1905	Device Revision or Replacement	A09	Output Problem	26/06/2014	
									A0907	No Device Output	26/06/2014	
							F21	Unexpected Deterioration	A09	Output Problem	26/06/2014	
									A0907	No Device Output	26/06/2014	
					CT1685	Cardiac pacemakers	F1905	Device Revision or Replacement	A09	Output Problem	26/06/2014	
									A0907	No Device Output	26/06/2014	
							F21	Unexpected Deterioration	A09	Output Problem	26/06/2014	
									A0907	No Device Output	26/06/2014	
					CT257	Implantable cardiac pacemakers	F1905	Device Revision or Replacement	A09	Output Problem	26/06/2014	
									A0907	No Device Output	26/06/2014	
							F21	Unexpected Deterioration	A09	Output Problem	26/06/2014	
									A0907	No Device Output	26/06/2014	
					CT2866	Implantable cardiac electrical stimulators	F1905	Device Revision or Replacement	A09	Output Problem	26/06/2014	
									A0907	No Device Output	26/06/2014	
							F21	Unexpected Deterioration	A09	Output Problem	26/06/2014	
									A0907	No Device Output	26/06/2014	
3266	2014	All other reportable incidents			CT257	Implantable cardiac pacemakers	F1905	Device Revision or Replacement	A1304	Telemetry Discrepancy		
									A26	Insufficient Device Problem Information		
3267	2014	All other reportable incidents			CT257	Implantable cardiac pacemakers	F1905	Device Revision or Replacement	A07	Electrical /Electronic Property Problem	01/07/2014	
									A0708	Power Problem	01/07/2014	
3268	2014				CT257	Implantable cardiac pacemakers	F21	Unexpected Deterioration	A26	Insufficient Device Problem Information		
3269	2014		47270	Cardiac resynchronization therapy implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	F1905	Device Revision or Replacement	A07	Electrical /Electronic Property Problem		
									A0708	Power Problem		
3270	2014	All other reportable incidents			CT257	Implantable cardiac pacemakers			A07	Electrical /Electronic Property Problem		
									A0709	Device Sensing Problem		
3271	2014				CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A07	Electrical /Electronic Property Problem		
									A0709	Device Sensing Problem		
3272	2013		47270	Cardiac resynchronization therapy implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	F1905	Device Revision or Replacement	A07	Electrical /Electronic Property Problem		
									A0709	Device Sensing Problem		
3273	2013		47270	Cardiac resynchronization therapy implantable defibrillator	CT2866	Implantable cardiac electrical stimulators			A26	Insufficient Device Problem Information		
3274	2013		47270	Cardiac resynchronization therapy implantable defibrillator	CT2866	Implantable cardiac electrical stimulators			A26	Insufficient Device Problem Information		
3275	2014		47270	Cardiac resynchronization therapy implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A07	Electrical /Electronic Property Problem		
									A0708	Power Problem		
3276	2013		47270	Cardiac resynchronization therapy implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E2104	Electric Shock	A09	Output Problem		
									A090404	Output Above Specifications		
							F14	Prolonged Episode of Care	A09	Output Problem		
									A090404	Output Above Specifications		
3277	2013				CT257	Implantable cardiac pacemakers	F1905	Device Revision or Replacement	A16	Protective Measures Problem		
									A1602	Fail-Safe Problem		
3278	2013				CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A07	Electrical /Electronic Property Problem		
									A0708	Power Problem		
3279	2014		47270	Cardiac resynchronization therapy implantable defibrillator	CT2866	Implantable cardiac electrical stimulators			A26	Insufficient Device Problem Information		
3280	2014		47270	Cardiac resynchronization therapy implantable defibrillator	CT2866	Implantable cardiac electrical stimulators			A26	Insufficient Device Problem Information		
3281	2014				CT257	Implantable cardiac pacemakers	F1905	Device Revision or Replacement	A0104	Migration or Expulsion of Device	27/07/2014	
3282	2013	All other reportable incidents			CT257	Implantable cardiac pacemakers	F1905	Device Revision or Replacement	A07	Electrical /Electronic Property Problem	07/10/2013	
									A0721	Circuit Failure	07/10/2013	
3283	2013		47270	Cardiac resynchronization therapy implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	F1905	Device Revision or Replacement	A07	Electrical /Electronic Property Problem		
									A0708	Power Problem		
3284	2014				CT257	Implantable cardiac pacemakers	E0402	Hypersensitivity/Allergic Reaction	A26	Insufficient Device Problem Information		
3285	2014		47270	Cardiac resynchronization therapy implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E0402	Hypersensitivity/Allergic Reaction	A26	Insufficient Device Problem Information		
3286	2013				CT257	Implantable cardiac pacemakers	E2401	Insufficient Information	A07	Electrical /Electronic Property Problem		
									A0708	Power Problem		
3287	2013				CT257	Implantable cardiac pacemakers	F1905	Device Revision or Replacement	A07	Electrical /Electronic Property Problem		
									A0708	Power Problem		
3288	2014		35852	Single-chamber implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	F15	Recognized Device or Procedural Complication	A26	Insufficient Device Problem Information		
3289	2013				CT257	Implantable cardiac pacemakers	F1905	Device Revision or Replacement	A07	Electrical /Electronic Property Problem		
									A0708	Power Problem		
3290	2013		47270	Cardiac resynchronization therapy implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E2401	Insufficient Information	A26	Insufficient Device Problem Information		
3291	2013		47270	Cardiac resynchronization therapy implantable defibrillator	CT2866	Implantable cardiac electrical stimulators			A26	Insufficient Device Problem Information		
3292	2013		47270	Cardiac resynchronization therapy implantable defibrillator	CT2866	Implantable cardiac electrical stimulators			A26	Insufficient Device Problem Information		
3293	2013		47270	Cardiac resynchronization therapy implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A07	Electrical /Electronic Property Problem		
									A0709	Device Sensing Problem		
3294	2013				CT257	Implantable cardiac pacemakers	E2104	Electric Shock	A09	Output Problem		
									A090404	Output Above Specifications		
3295	2014		47270	Cardiac resynchronization therapy implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E0402	Hypersensitivity/Allergic Reaction	A26	Insufficient Device Problem Information		

3296	2014		47270	Cardiac resynchronization therapy implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E0402	Hypersensitivity/Allergic Reaction	A26	Insufficient Device Problem Information		
3297	2013				CT257	Implantable cardiac pacemakers	E2401	Insufficient Information	A13	Communication or Transmission Problem		
									A26	Insufficient Device Problem Information		
3298	2013	All other reportable incidents			CT257	Implantable cardiac pacemakers	F1905	Device Revision or Replacement	A07	Electrical /Electronic Property Problem	02/10/2013	
									A0708	Power Problem	02/10/2013	
3299	2013	All other reportable incidents			CT257	Implantable cardiac pacemakers	F1905	Device Revision or Replacement	A07	Electrical /Electronic Property Problem	07/10/2013	
									A0708	Power Problem	07/10/2013	
3300	2014	All other reportable incidents	35223	Endocardial/interventricular septal pacing lead	CT1684	Cardiac pacemakers and associated devices	F1905	Device Revision or Replacement	A0104	Migration or Expulsion of Device	09/07/2014	
3301	2014	All other reportable incidents	47270	Cardiac resynchronization therapy implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	F1905	Device Revision or Replacement	A07	Electrical /Electronic Property Problem	18/07/2014	
									A0708	Power Problem	18/07/2014	
3302	2014		35852	Single-chamber implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	F05	Delay to Treatment/ Therapy	A07	Electrical /Electronic Property Problem	01/08/2014	
									A0709	Device Sensing Problem	01/08/2014	
							F1905	Device Revision or Replacement	A07	Electrical /Electronic Property Problem	01/08/2014	
									A0709	Device Sensing Problem	01/08/2014	
3303	2014				CT257	Implantable cardiac pacemakers			A26	Insufficient Device Problem Information		
3304	2014				CT257	Implantable cardiac pacemakers	F1905	Device Revision or Replacement	A09	Output Problem	05/08/2014	
									A090404	Output Above Specifications	05/08/2014	
3305	2013				CT257	Implantable cardiac pacemakers	F1905	Device Revision or Replacement	A09	Output Problem	24/10/2013	
									A0907	No Device Output	24/10/2013	
									A13	Communication or Transmission Problem	24/10/2013	
									A26	Insufficient Device Problem Information	24/10/2013	
3306	2014	All other reportable incidents			CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A1501	Activation Problem		
									A23	Use of Device Problem		
3307	2014				CT257	Implantable cardiac pacemakers	E0402	Hypersensitivity/Allergic Reaction	A26	Insufficient Device Problem Information		
3308	2014				CT257	Implantable cardiac pacemakers	F21	Unexpected Deterioration	A26	Insufficient Device Problem Information		
3309	2013	All other reportable incidents			CT257	Implantable cardiac pacemakers	F1905	Device Revision or Replacement	A09	Output Problem		
									A0907	No Device Output		
3310	2014		35852	Single-chamber implantable defibrillator	CT2866	Implantable cardiac electrical stimulators			A07	Electrical /Electronic Property Problem		
									A0709	Device Sensing Problem		
3311	2013		47270	Cardiac resynchronization therapy implantable defibrillator	CT2866	Implantable cardiac electrical stimulators			A26	Insufficient Device Problem Information		
3312	2013		47270	Cardiac resynchronization therapy implantable defibrillator	CT2866	Implantable cardiac electrical stimulators			A26	Insufficient Device Problem Information		
3313	2013		47270	Cardiac resynchronization therapy implantable defibrillator	CT2866	Implantable cardiac electrical stimulators			A26	Insufficient Device Problem Information		
3314	2013		47270	Cardiac resynchronization therapy implantable defibrillator	CT2866	Implantable cardiac electrical stimulators			A26	Insufficient Device Problem Information		
3315	2013		47270	Cardiac resynchronization therapy implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	F21	Unexpected Deterioration	A26	Insufficient Device Problem Information		
3316	2013		47270	Cardiac resynchronization therapy implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A26	Insufficient Device Problem Information		
3317	2013				CT257	Implantable cardiac pacemakers			A1104	Computer Operating System Problem		
3318	2013		47270	Cardiac resynchronization therapy implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	F15	Recognized Device or Procedural Complication	A26	Insufficient Device Problem Information		
3319	2013		35852	Single-chamber implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A07	Electrical /Electronic Property Problem		
									A0708	Power Problem		
3320	2015		47270	Cardiac resynchronization therapy implantable defibrillator	CT2866	Implantable cardiac electrical stimulators			A26	Insufficient Device Problem Information		
3321	2015		47270	Cardiac resynchronization therapy implantable defibrillator	CT2866	Implantable cardiac electrical stimulators			A26	Insufficient Device Problem Information		
3322	2015		47270	Cardiac resynchronization therapy implantable defibrillator	CT2866	Implantable cardiac electrical stimulators			A26	Insufficient Device Problem Information		
3323	2015		47270	Cardiac resynchronization therapy implantable defibrillator	CT2866	Implantable cardiac electrical stimulators			A26	Insufficient Device Problem Information		
3324	2016	TBD	35223	Endocardial/interventricular septal pacing lead	CT1684	Cardiac pacemakers and associated devices			A26	Insufficient Device Problem Information		
3325	2015				CT257	Implantable cardiac pacemakers			A26	Insufficient Device Problem Information		
3326	2015		47270	Cardiac resynchronization therapy implantable defibrillator	CT2866	Implantable cardiac electrical stimulators			A26	Insufficient Device Problem Information		
3327	2016	TBD			CT257	Implantable cardiac pacemakers			A26	Insufficient Device Problem Information		
3328	2015		47270	Cardiac resynchronization therapy implantable defibrillator	CT2866	Implantable cardiac electrical stimulators			A26	Insufficient Device Problem Information		
3329	2015				CT257	Implantable cardiac pacemakers			A26	Insufficient Device Problem Information		
3330	2016	All other reportable incidents			CT1684	Cardiac pacemakers and associated devices			A09	Output Problem		
									A0908	Incorrect, Inadequate or Imprecise Result or Readings		
3331	2016	All other reportable incidents			CT1684	Cardiac pacemakers and associated devices			A09	Output Problem		
									A0908	Incorrect, Inadequate or Imprecise Result or Readings		
3332	2016	TBD	47270	Cardiac resynchronization therapy implantable defibrillator	CT2866	Implantable cardiac electrical stimulators			A07	Electrical /Electronic Property Problem		
									A0709	Device Sensing Problem		
3333	2016	All other reportable incidents	47263	Cardiac resynchronization therapy implantable pacemaker	CT1684	Cardiac pacemakers and associated devices			A07	Electrical /Electronic Property Problem	16/12/2015	
									A0708	Power Problem	16/12/2015	
					CT1685	Cardiac pacemakers			A07	Electrical /Electronic Property Problem	16/12/2015	
									A0708	Power Problem	16/12/2015	
					CT257	Implantable cardiac pacemakers			A07	Electrical /Electronic Property Problem	16/12/2015	
									A0708	Power Problem	16/12/2015	
					CT2866	Implantable cardiac electrical stimulators			A07	Electrical /Electronic Property Problem	16/12/2015	
									A0708	Power Problem	16/12/2015	
3334	2016	All other reportable incidents	47263	Cardiac resynchronization therapy implantable pacemaker	CT1684	Cardiac pacemakers and associated devices			A07	Electrical /Electronic Property Problem	31/05/2016	
									A0708	Power Problem	31/05/2016	
					CT1685	Cardiac pacemakers			A07	Electrical /Electronic Property Problem	31/05/2016	
									A0708	Power Problem	31/05/2016	
					CT257	Implantable cardiac pacemakers			A07	Electrical /Electronic Property Problem	31/05/2016	
									A0708	Power Problem	31/05/2016	
					CT2866	Implantable cardiac electrical stimulators			A07	Electrical /Electronic Property Problem	31/05/2016	
									A0708	Power Problem	31/05/2016	
3335	2015	All other reportable incidents	47270	Cardiac resynchronization therapy implantable defibrillator	CT2866	Implantable cardiac electrical stimulators			A07	Electrical /Electronic Property Problem		
									A0709	Device Sensing Problem		
3336	2015	All other reportable incidents			CT257	Implantable cardiac pacemakers			A07	Electrical /Electronic Property Problem		

									A0708	Power Problem		
									A16	Protective Measures Problem		
									A1602	Fail-Safe Problem		
3337	2015				CT1684	Cardiac pacemakers and associated devices			A15	Activation, Positioning or Separation Problem		
									A150101	Activation Failure		
									A150102	Difficult or Delayed Activation		
									A150103	Premature Activation		
3338	2015				CT257	Implantable cardiac pacemakers			A01	Patient Device Interaction Problem		
									A17	Compatibility Problem		
3339	2016	TBD			CT1684	Cardiac pacemakers and associated devices			A09	Output Problem		
									A0908	Incorrect, Inadequate or Imprecise Result or Readings		
3340	2015		47270	Cardiac resynchronization therapy implantable defibrillator	CT2866	Implantable cardiac electrical stimulators			A26	Insufficient Device Problem Information		
3341	2015		47270	Cardiac resynchronization therapy implantable defibrillator	CT2866	Implantable cardiac electrical stimulators			A07	Electrical /Electronic Property Problem		
									A0708	Power Problem		
3342	2016	TBD	35223	Endocardial/interventricular septal pacing lead	CT1684	Cardiac pacemakers and associated devices			A26	Insufficient Device Problem Information		
3343	2016	TBD	35223	Endocardial/interventricular septal pacing lead	CT1684	Cardiac pacemakers and associated devices			A26	Insufficient Device Problem Information		
3344	2016	TBD			CT257	Implantable cardiac pacemakers			A07	Electrical /Electronic Property Problem		
									A0708	Power Problem		
3345	2015		47270	Cardiac resynchronization therapy implantable defibrillator	CT2866	Implantable cardiac electrical stimulators			A07	Electrical /Electronic Property Problem		
									A0708	Power Problem		
3346	2016	All other reportable incidents	47263	Cardiac resynchronization therapy implantable pacemaker	CT1684	Cardiac pacemakers and associated devices			A07	Electrical /Electronic Property Problem	07/12/2015	
									A0708	Power Problem	07/12/2015	
					CT1685	Cardiac pacemakers			A07	Electrical /Electronic Property Problem	07/12/2015	
									A0708	Power Problem	07/12/2015	
					CT257	Implantable cardiac pacemakers			A07	Electrical /Electronic Property Problem	07/12/2015	
									A0708	Power Problem	07/12/2015	
					CT2866	Implantable cardiac electrical stimulators			A07	Electrical /Electronic Property Problem	07/12/2015	
									A0708	Power Problem	07/12/2015	
3347	2016	All other reportable incidents	47263	Cardiac resynchronization therapy implantable pacemaker	CT1684	Cardiac pacemakers and associated devices			A07	Electrical /Electronic Property Problem	08/01/2016	
									A0708	Power Problem	08/01/2016	
					CT1685	Cardiac pacemakers			A07	Electrical /Electronic Property Problem	08/01/2016	
									A0708	Power Problem	08/01/2016	
					CT257	Implantable cardiac pacemakers			A07	Electrical /Electronic Property Problem	08/01/2016	
									A0708	Power Problem	08/01/2016	
					CT2866	Implantable cardiac electrical stimulators			A07	Electrical /Electronic Property Problem	08/01/2016	
									A0708	Power Problem	08/01/2016	
3348	2016	All other reportable incidents	47263	Cardiac resynchronization therapy implantable pacemaker	CT1684	Cardiac pacemakers and associated devices			A07	Electrical /Electronic Property Problem	08/01/2016	
									A0708	Power Problem	08/01/2016	
					CT1685	Cardiac pacemakers			A07	Electrical /Electronic Property Problem	08/01/2016	
									A0708	Power Problem	08/01/2016	
					CT257	Implantable cardiac pacemakers			A07	Electrical /Electronic Property Problem	08/01/2016	
									A0708	Power Problem	08/01/2016	
					CT2866	Implantable cardiac electrical stimulators			A07	Electrical /Electronic Property Problem	08/01/2016	
									A0708	Power Problem	08/01/2016	
3349	2016	All other reportable incidents	47263	Cardiac resynchronization therapy implantable pacemaker	CT1684	Cardiac pacemakers and associated devices			A07	Electrical /Electronic Property Problem	08/01/2016	
									A0708	Power Problem	08/01/2016	
					CT1685	Cardiac pacemakers			A07	Electrical /Electronic Property Problem	08/01/2016	
									A0708	Power Problem	08/01/2016	
					CT257	Implantable cardiac pacemakers			A07	Electrical /Electronic Property Problem	08/01/2016	
									A0708	Power Problem	08/01/2016	
					CT2866	Implantable cardiac electrical stimulators			A07	Electrical /Electronic Property Problem	08/01/2016	
									A0708	Power Problem	08/01/2016	
3350	2016	All other reportable incidents	47263	Cardiac resynchronization therapy implantable pacemaker	CT1684	Cardiac pacemakers and associated devices			A07	Electrical /Electronic Property Problem		
									A0708	Power Problem		
					CT1685	Cardiac pacemakers			A07	Electrical /Electronic Property Problem		
									A0708	Power Problem		
					CT257	Implantable cardiac pacemakers			A07	Electrical /Electronic Property Problem		
									A0708	Power Problem		
					CT2866	Implantable cardiac electrical stimulators			A07	Electrical /Electronic Property Problem		
									A0708	Power Problem		
3351	2016	All other reportable incidents	47263	Cardiac resynchronization therapy implantable pacemaker	CT1684	Cardiac pacemakers and associated devices			A07	Electrical /Electronic Property Problem	20/01/2016	
									A0708	Power Problem	20/01/2016	
					CT1685	Cardiac pacemakers			A07	Electrical /Electronic Property Problem	20/01/2016	
									A0708	Power Problem	20/01/2016	
					CT257	Implantable cardiac pacemakers			A07	Electrical /Electronic Property Problem	20/01/2016	
									A0708	Power Problem	20/01/2016	
					CT2866	Implantable cardiac electrical stimulators			A07	Electrical /Electronic Property Problem	20/01/2016	
									A0708	Power Problem	20/01/2016	
3352	2016	All other reportable incidents	47263	Cardiac resynchronization therapy implantable pacemaker	CT1684	Cardiac pacemakers and associated devices			A07	Electrical /Electronic Property Problem		
									A0708	Power Problem		
					CT1685	Cardiac pacemakers			A07	Electrical /Electronic Property Problem		
									A0708	Power Problem		
					CT257	Implantable cardiac pacemakers			A07	Electrical /Electronic Property Problem		
									A0708	Power Problem		
					CT2866	Implantable cardiac electrical stimulators			A07	Electrical /Electronic Property Problem		
									A0708	Power Problem		
3353	2015				CT257	Implantable cardiac pacemakers			A07	Electrical /Electronic Property Problem		
									A0709	Device Sensing Problem		
3354	2016	All other reportable incidents	47263	Cardiac resynchronization therapy implantable pacemaker	CT1684	Cardiac pacemakers and associated devices			A07	Electrical /Electronic Property Problem		
									A0708	Power Problem		
					CT1685	Cardiac pacemakers			A07	Electrical /Electronic Property Problem		
									A0708	Power Problem		
					CT257	Implantable cardiac pacemakers			A07	Electrical /Electronic Property Problem		
									A0708	Power Problem		
					CT2866	Implantable cardiac electrical stimulators			A07	Electrical /Electronic Property Problem		

									A0708	Power Problem		
3355	2016	All other reportable incidents	47263	Cardiac resynchronization therapy implantable pacemaker	CT1684	Cardiac pacemakers and associated devices			A07	Electrical /Electronic Property Problem		
									A0708	Power Problem		
					CT1685	Cardiac pacemakers			A07	Electrical /Electronic Property Problem		
									A0708	Power Problem		
					CT257	Implantable cardiac pacemakers			A07	Electrical /Electronic Property Problem		
									A0708	Power Problem		
					CT2866	Implantable cardiac electrical stimulators			A07	Electrical /Electronic Property Problem		
									A0708	Power Problem		
3356	2015				CT257	Implantable cardiac pacemakers			A26	Insufficient Device Problem Information		
3357	2016	TBD			CT257	Implantable cardiac pacemakers			A07	Electrical /Electronic Property Problem		
									A0708	Power Problem		
3358	2015	All other reportable incidents	47270	Cardiac resynchronization therapy implantable defibrillator	CT2866	Implantable cardiac electrical stimulators			A07	Electrical /Electronic Property Problem	18/12/2014	
									A0708	Power Problem	18/12/2014	
3359	2015	All other reportable incidents			CT257	Implantable cardiac pacemakers			A07	Electrical /Electronic Property Problem	08/12/2014	
									A0708	Power Problem	08/12/2014	
3360	2015		47270	Cardiac resynchronization therapy implantable defibrillator	CT2866	Implantable cardiac electrical stimulators			A07	Electrical /Electronic Property Problem		
									A0708	Power Problem		
3361	2015		47270	Cardiac resynchronization therapy implantable defibrillator	CT2866	Implantable cardiac electrical stimulators			A07	Electrical /Electronic Property Problem		
									A0708	Power Problem		
3362	2016	All other reportable incidents	35039	Epicardial pacing lead	CT1684	Cardiac pacemakers and associated devices			A07	Electrical /Electronic Property Problem		
									A0709	Device Sensing Problem		
3363	2016	All other reportable incidents			CT257	Implantable cardiac pacemakers			A07	Electrical /Electronic Property Problem	20/07/2016	
									A0708	Power Problem	20/07/2016	
3364	2016	TBD	47270	Cardiac resynchronization therapy implantable defibrillator	CT2866	Implantable cardiac electrical stimulators			A07	Electrical /Electronic Property Problem		
									A0709	Device Sensing Problem		
3365	2015	All other reportable incidents			CT257	Implantable cardiac pacemakers			A07	Electrical /Electronic Property Problem	14/05/2015	
									A0708	Power Problem	14/05/2015	
3366	2015		47270	Cardiac resynchronization therapy implantable defibrillator	CT2866	Implantable cardiac electrical stimulators			A26	Insufficient Device Problem Information		
3367	2015	All other reportable incidents	47270	Cardiac resynchronization therapy implantable defibrillator	CT2866	Implantable cardiac electrical stimulators			A07	Electrical /Electronic Property Problem		
									A0708	Power Problem		
3368	2016	TBD			CT257	Implantable cardiac pacemakers			A07	Electrical /Electronic Property Problem		
									A0709	Device Sensing Problem		
3369	2016	TBD			CT257	Implantable cardiac pacemakers			A07	Electrical /Electronic Property Problem		
									A0708	Power Problem		
3370	2015				CT257	Implantable cardiac pacemakers			A26	Insufficient Device Problem Information		
3371	2015				CT257	Implantable cardiac pacemakers			A07	Electrical /Electronic Property Problem		
									A0708	Power Problem		
3372	2016	All other reportable incidents			CT1684	Cardiac pacemakers and associated devices			A09	Output Problem		
									A0907	No Device Output		
3373	2016	TBD	47270	Cardiac resynchronization therapy implantable defibrillator	CT2866	Implantable cardiac electrical stimulators			A07	Electrical /Electronic Property Problem		
									A0709	Device Sensing Problem		
3374	2016	TBD	47270	Cardiac resynchronization therapy implantable defibrillator	CT2866	Implantable cardiac electrical stimulators			A07	Electrical /Electronic Property Problem		
									A0708	Power Problem		
3375	2015	All other reportable incidents			CT257	Implantable cardiac pacemakers			A04	Material Integrity Problem		
									A0404	Crack		
3376	2015		47270	Cardiac resynchronization therapy implantable defibrillator	CT2866	Implantable cardiac electrical stimulators			A24	No Apparent Device Problem		
									A26	Insufficient Device Problem Information		
3377	2015	All other reportable incidents			CT257	Implantable cardiac pacemakers			A07	Electrical /Electronic Property Problem	05/10/2015	
									A0708	Power Problem	05/10/2015	
3378	2016	TBD	47270	Cardiac resynchronization therapy implantable defibrillator	CT2866	Implantable cardiac electrical stimulators			A11	Computer Software Problem		
									A1102	Application Program Problem		
3379	2015		47270	Cardiac resynchronization therapy implantable defibrillator	CT2866	Implantable cardiac electrical stimulators			A26	Insufficient Device Problem Information		
3380	2016	TBD	35223	Endocardial/interventricular septal pacing lead	CT1684	Cardiac pacemakers and associated devices			A26	Insufficient Device Problem Information		
3381	2016	TBD	47270	Cardiac resynchronization therapy implantable defibrillator	CT2866	Implantable cardiac electrical stimulators			A26	Insufficient Device Problem Information		
3382	2016	All other reportable incidents			CT257	Implantable cardiac pacemakers			A07	Electrical /Electronic Property Problem		
									A0709	Device Sensing Problem		
									A1304	Telemetry Discrepancy		
									A26	Insufficient Device Problem Information		
3383	2015				CT257	Implantable cardiac pacemakers			A07	Electrical /Electronic Property Problem		
									A0708	Power Problem		
3384	2015				CT257	Implantable cardiac pacemakers			A07	Electrical /Electronic Property Problem		
									A0708	Power Problem		
3385	2016	TBD			CT257	Implantable cardiac pacemakers			A07	Electrical /Electronic Property Problem		
									A0708	Power Problem		
3386	2016	TBD	35223	Endocardial/interventricular septal pacing lead	CT1684	Cardiac pacemakers and associated devices			A26	Insufficient Device Problem Information		
3387	2015				CT257	Implantable cardiac pacemakers			A26	Insufficient Device Problem Information		
3388	2015	All other reportable incidents	35223	Endocardial/interventricular septal pacing lead	CT1684	Cardiac pacemakers and associated devices			A07	Electrical /Electronic Property Problem		
									A0709	Device Sensing Problem		
3389	2015		47270	Cardiac resynchronization therapy implantable defibrillator	CT2866	Implantable cardiac electrical stimulators			A26	Insufficient Device Problem Information		
3390	2015		47270	Cardiac resynchronization therapy implantable defibrillator	CT2866	Implantable cardiac electrical stimulators			A26	Insufficient Device Problem Information		
3391	2015	All other reportable incidents			CT257	Implantable cardiac pacemakers			A07	Electrical /Electronic Property Problem		
									A0708	Power Problem		
3392	2016	TBD			CT257	Implantable cardiac pacemakers			A07	Electrical /Electronic Property Problem		
									A0708	Power Problem		
3393	2015	All other reportable incidents			CT1684	Cardiac pacemakers and associated devices			A07	Electrical /Electronic Property Problem		
									A0708	Power Problem		
3394	2015	All other reportable incidents			CT1684	Cardiac pacemakers and associated devices			A07	Electrical /Electronic Property Problem		
									A0708	Power Problem		
3395	2015				CT257	Implantable cardiac pacemakers			A26	Insufficient Device Problem Information		
3396	2015		47270	Cardiac resynchronization therapy implantable defibrillator	CT2866	Implantable cardiac electrical stimulators			A07	Electrical /Electronic Property Problem		
									A0709	Device Sensing Problem		

3397	2015		47270	Cardiac resynchronization therapy implantable defibrillator	CT2866	Implantable cardiac electrical stimulators			A07	Electrical /Electronic Property Problem		
									A0709	Device Sensing Problem		
3398	2016	All other reportable incidents			CT257	Implantable cardiac pacemakers			A07	Electrical /Electronic Property Problem	07/03/2016	
									A0709	Device Sensing Problem	07/03/2016	
3399	2015				CT257	Implantable cardiac pacemakers			A26	Insufficient Device Problem Information		
3400	2016	All other reportable incidents	47263	Cardiac resynchronization therapy implantable pacemaker	CT1684	Cardiac pacemakers and associated devices			A07	Electrical /Electronic Property Problem	23/02/2016	
									A0709	Device Sensing Problem	23/02/2016	
					CT1685	Cardiac pacemakers			A07	Electrical /Electronic Property Problem	23/02/2016	
									A0709	Device Sensing Problem	23/02/2016	
					CT257	Implantable cardiac pacemakers			A07	Electrical /Electronic Property Problem	23/02/2016	
									A0709	Device Sensing Problem	23/02/2016	
					CT2866	Implantable cardiac electrical stimulators			A07	Electrical /Electronic Property Problem	23/02/2016	
									A0709	Device Sensing Problem	23/02/2016	
3401	2016	All other reportable incidents	47263	Cardiac resynchronization therapy implantable pacemaker	CT1684	Cardiac pacemakers and associated devices			A07	Electrical /Electronic Property Problem	11/03/2016	
									A0708	Power Problem	11/03/2016	
					CT1685	Cardiac pacemakers			A07	Electrical /Electronic Property Problem	11/03/2016	
									A0708	Power Problem	11/03/2016	
					CT257	Implantable cardiac pacemakers			A07	Electrical /Electronic Property Problem	11/03/2016	
									A0708	Power Problem	11/03/2016	
					CT2866	Implantable cardiac electrical stimulators			A07	Electrical /Electronic Property Problem	11/03/2016	
									A0708	Power Problem	11/03/2016	
3402	2015		47270	Cardiac resynchronization therapy implantable defibrillator	CT2866	Implantable cardiac electrical stimulators			A01	Patient Device Interaction Problem		
									A17	Compatibility Problem		
3403	2016	Death	47270	Cardiac resynchronization therapy implantable defibrillator	CT2866	Implantable cardiac electrical stimulators			A26	Insufficient Device Problem Information		
3404	2016	All other reportable incidents			CT257	Implantable cardiac pacemakers			A07	Electrical /Electronic Property Problem	13/07/2016	
									A0709	Device Sensing Problem	13/07/2016	
3405	2015	All other reportable incidents			CT257	Implantable cardiac pacemakers			A07	Electrical /Electronic Property Problem		
									A0708	Power Problem		
3406	2015		35852	Single-chamber implantable defibrillator	CT2866	Implantable cardiac electrical stimulators			A01	Patient Device Interaction Problem		
									A17	Compatibility Problem		
3407	2016	TBD			CT257	Implantable cardiac pacemakers			A07	Electrical /Electronic Property Problem		
									A0708	Power Problem		
3408	2015				CT257	Implantable cardiac pacemakers			A07	Electrical /Electronic Property Problem		
									A0708	Power Problem		
3409	2015				CT257	Implantable cardiac pacemakers			A1104	Computer Operating System Problem		
3410	2015				CT257	Implantable cardiac pacemakers			A07	Electrical /Electronic Property Problem		
									A0708	Power Problem		
3411	2016	All other reportable incidents	47263	Cardiac resynchronization therapy implantable pacemaker	CT1684	Cardiac pacemakers and associated devices			A07	Electrical /Electronic Property Problem	07/12/2015	
									A0708	Power Problem	07/12/2015	
					CT1685	Cardiac pacemakers			A07	Electrical /Electronic Property Problem	07/12/2015	
									A0708	Power Problem	07/12/2015	
					CT257	Implantable cardiac pacemakers			A07	Electrical /Electronic Property Problem	07/12/2015	
									A0708	Power Problem	07/12/2015	
					CT2866	Implantable cardiac electrical stimulators			A07	Electrical /Electronic Property Problem	07/12/2015	
									A0708	Power Problem	07/12/2015	
3412	2015	All other reportable incidents			CT1684	Cardiac pacemakers and associated devices			A07	Electrical /Electronic Property Problem		
									A0708	Power Problem		
3413	2015				CT257	Implantable cardiac pacemakers			A07	Electrical /Electronic Property Problem		
									A0709	Device Sensing Problem		
3414	2016	TBD	47270	Cardiac resynchronization therapy implantable defibrillator	CT2866	Implantable cardiac electrical stimulators			A09	Output Problem		
									A0907	No Device Output		
3415	2015				CT257	Implantable cardiac pacemakers			A26	Insufficient Device Problem Information		
3416	2015				CT257	Implantable cardiac pacemakers			A07	Electrical /Electronic Property Problem		
									A0708	Power Problem		
3417	2015				CT257	Implantable cardiac pacemakers			A07	Electrical /Electronic Property Problem		
									A0708	Power Problem		
3418	2015		47270	Cardiac resynchronization therapy implantable defibrillator	CT2866	Implantable cardiac electrical stimulators			A07	Electrical /Electronic Property Problem		
									A0708	Power Problem		
3419	2015	All other reportable incidents			CT1684	Cardiac pacemakers and associated devices			A07	Electrical /Electronic Property Problem		
									A0708	Power Problem		
3420	2016	All other reportable incidents	47263	Cardiac resynchronization therapy implantable pacemaker	CT1684	Cardiac pacemakers and associated devices			A07	Electrical /Electronic Property Problem	16/05/2016	
									A0708	Power Problem	16/05/2016	
					CT1685	Cardiac pacemakers			A07	Electrical /Electronic Property Problem	16/05/2016	
									A0708	Power Problem	16/05/2016	
					CT257	Implantable cardiac pacemakers			A07	Electrical /Electronic Property Problem	16/05/2016	
									A0708	Power Problem	16/05/2016	
					CT2866	Implantable cardiac electrical stimulators			A07	Electrical /Electronic Property Problem	16/05/2016	
									A0708	Power Problem	16/05/2016	
3421	2016	All other reportable incidents			CT257	Implantable cardiac pacemakers			A07	Electrical /Electronic Property Problem	07/03/2016	
									A0708	Power Problem	07/03/2016	
3422	2015		47270	Cardiac resynchronization therapy implantable defibrillator	CT2866	Implantable cardiac electrical stimulators			A26	Insufficient Device Problem Information		
3423	2016	All other reportable incidents	47270	Cardiac resynchronization therapy implantable defibrillator	CT2866	Implantable cardiac electrical stimulators			A16	Protective Measures Problem		
									A1602	Fail-Safe Problem		
3424	2016	TBD	35223	Endocardial/interventricular septal pacing lead	CT1684	Cardiac pacemakers and associated devices			A26	Insufficient Device Problem Information		
3425	2015	All other reportable incidents			CT257	Implantable cardiac pacemakers			A13	Communication or Transmission Problem		
									A26	Insufficient Device Problem Information		
3426	2015		47270	Cardiac resynchronization therapy implantable defibrillator	CT2866	Implantable cardiac electrical stimulators			A26	Insufficient Device Problem Information		
3427	2016	TBD			CT257	Implantable cardiac pacemakers			A07	Electrical /Electronic Property Problem		
									A0709	Device Sensing Problem		
3428	2015				CT257	Implantable cardiac pacemakers			A07	Electrical /Electronic Property Problem		
									A0708	Power Problem		
3429	2015				CT257	Implantable cardiac pacemakers			A07	Electrical /Electronic Property Problem		
									A0709	Device Sensing Problem		
3430	2015				CT257	Implantable cardiac pacemakers			A07	Electrical /Electronic Property Problem		
									A0709	Device Sensing Problem		

3431	2016	TBD	47270	Cardiac resynchronization therapy implantable defibrillator	CT2866	Implantable cardiac electrical stimulators			A26	Insufficient Device Problem Information		
3432	2016	All other reportable incidents			CT257	Implantable cardiac pacemakers			A07	Electrical /Electronic Property Problem	19/04/2016	
									A0708	Power Problem	19/04/2016	
3433	2016	TBD	35223	Endocardial/interventricular septal pacing lead	CT1684	Cardiac pacemakers and associated devices			A26	Insufficient Device Problem Information		
3434	2016	TBD	47270	Cardiac resynchronization therapy implantable defibrillator	CT2866	Implantable cardiac electrical stimulators			A07	Electrical /Electronic Property Problem		
									A0708	Power Problem		
3435	2016	TBD			CT257	Implantable cardiac pacemakers			A09	Output Problem		
									A0907	No Device Output		
3436	2015				CT257	Implantable cardiac pacemakers			A26	Insufficient Device Problem Information		
3437	2016	All other reportable incidents	47263	Cardiac resynchronization therapy implantable pacemaker	CT1684	Cardiac pacemakers and associated devices			A07	Electrical /Electronic Property Problem	22/12/2015	
									A0708	Power Problem	22/12/2015	
					CT1685	Cardiac pacemakers			A07	Electrical /Electronic Property Problem	22/12/2015	
									A0708	Power Problem	22/12/2015	
					CT257	Implantable cardiac pacemakers			A07	Electrical /Electronic Property Problem	22/12/2015	
									A0708	Power Problem	22/12/2015	
					CT2866	Implantable cardiac electrical stimulators			A07	Electrical /Electronic Property Problem	22/12/2015	
									A0708	Power Problem	22/12/2015	
3438	2015	All other reportable incidents			CT257	Implantable cardiac pacemakers			A0104	Migration or Expulsion of Device	09/01/2015	
3439	2015	All other reportable incidents			CT257	Implantable cardiac pacemakers			A07	Electrical /Electronic Property Problem		
									A0708	Power Problem		
3440	2016	All other reportable incidents			CT257	Implantable cardiac pacemakers			A13	Communication or Transmission Problem		
									A26	Insufficient Device Problem Information		
3441	2015				CT257	Implantable cardiac pacemakers			A26	Insufficient Device Problem Information		
3442	2015		47270	Cardiac resynchronization therapy implantable defibrillator	CT2866	Implantable cardiac electrical stimulators			A26	Insufficient Device Problem Information		
3443	2016	All other reportable incidents	47270	Cardiac resynchronization therapy implantable defibrillator	CT2866	Implantable cardiac electrical stimulators			A07	Electrical /Electronic Property Problem	06/09/2016	
									A0708	Power Problem	06/09/2016	
3444	2015	All other reportable incidents	47270	Cardiac resynchronization therapy implantable defibrillator	CT2866	Implantable cardiac electrical stimulators			A07	Electrical /Electronic Property Problem		
									A0708	Power Problem		
3445	2015				CT257	Implantable cardiac pacemakers			A26	Insufficient Device Problem Information		
3446	2015		47270	Cardiac resynchronization therapy implantable defibrillator	CT2866	Implantable cardiac electrical stimulators			A26	Insufficient Device Problem Information		
3447	2016	TBD	35852	Single-chamber implantable defibrillator	CT2866	Implantable cardiac electrical stimulators			A07	Electrical /Electronic Property Problem	14/06/2018	
									A0709	Device Sensing Problem	14/06/2018	
3448	2015		47270	Cardiac resynchronization therapy implantable defibrillator	CT2866	Implantable cardiac electrical stimulators			A26	Insufficient Device Problem Information		
3449	2015				CT257	Implantable cardiac pacemakers			A26	Insufficient Device Problem Information		
3450	2016	All other reportable incidents	47263	Cardiac resynchronization therapy implantable pacemaker	CT1684	Cardiac pacemakers and associated devices			A07	Electrical /Electronic Property Problem	17/03/2016	
									A0708	Power Problem	17/03/2016	
					CT1685	Cardiac pacemakers			A07	Electrical /Electronic Property Problem	17/03/2016	
									A0708	Power Problem	17/03/2016	
					CT257	Implantable cardiac pacemakers			A07	Electrical /Electronic Property Problem	17/03/2016	
									A0708	Power Problem	17/03/2016	
					CT2866	Implantable cardiac electrical stimulators			A07	Electrical /Electronic Property Problem	17/03/2016	
									A0708	Power Problem	17/03/2016	
3451	2015		47270	Cardiac resynchronization therapy implantable defibrillator	CT2866	Implantable cardiac electrical stimulators			A07	Electrical /Electronic Property Problem		
									A0708	Power Problem		
3452	2016	All other reportable incidents			CT257	Implantable cardiac pacemakers			A07	Electrical /Electronic Property Problem	23/11/2015	
									A0708	Power Problem	23/11/2015	
3453	2015		47270	Cardiac resynchronization therapy implantable defibrillator	CT2866	Implantable cardiac electrical stimulators			A07	Electrical /Electronic Property Problem		
									A0708	Power Problem		
3454	2016	TBD	47270	Cardiac resynchronization therapy implantable defibrillator	CT2866	Implantable cardiac electrical stimulators			A07	Electrical /Electronic Property Problem		
									A0709	Device Sensing Problem		
3455	2015	All other reportable incidents			CT257	Implantable cardiac pacemakers			A07	Electrical /Electronic Property Problem	21/10/2015	
									A0708	Power Problem	21/10/2015	
3456	2015				CT257	Implantable cardiac pacemakers			A07	Electrical /Electronic Property Problem		
									A0709	Device Sensing Problem		
3457	2015	All other reportable incidents	47270	Cardiac resynchronization therapy implantable defibrillator	CT2866	Implantable cardiac electrical stimulators			A07	Electrical /Electronic Property Problem	02/06/2015	
									A0708	Power Problem	02/06/2015	
3458	2015		47270	Cardiac resynchronization therapy implantable defibrillator	CT2866	Implantable cardiac electrical stimulators			A26	Insufficient Device Problem Information		
3459	2015		47270	Cardiac resynchronization therapy implantable defibrillator	CT2866	Implantable cardiac electrical stimulators			A26	Insufficient Device Problem Information		
3460	2015				CT257	Implantable cardiac pacemakers			A26	Insufficient Device Problem Information		
3461	2016	TBD			CT257	Implantable cardiac pacemakers			A09	Output Problem	01/09/2016	
									A0907	No Device Output	01/09/2016	
3462	2015		47270	Cardiac resynchronization therapy implantable defibrillator	CT2866	Implantable cardiac electrical stimulators			A26	Insufficient Device Problem Information		
3463	2016	TBD			CT257	Implantable cardiac pacemakers			A07	Electrical /Electronic Property Problem		
									A0708	Power Problem		
3464	2016	TBD	47270	Cardiac resynchronization therapy implantable defibrillator	CT2866	Implantable cardiac electrical stimulators			A07	Electrical /Electronic Property Problem		
									A0709	Device Sensing Problem		
3465	2016	TBD	35223	Endocardial/interventricular septal pacing lead	CT1684	Cardiac pacemakers and associated devices			A26	Insufficient Device Problem Information		
3466	2015		47270	Cardiac resynchronization therapy implantable defibrillator	CT2866	Implantable cardiac electrical stimulators			A26	Insufficient Device Problem Information		
3467	2015		47270	Cardiac resynchronization therapy implantable defibrillator	CT2866	Implantable cardiac electrical stimulators			A09	Output Problem		
									A090404	Output Above Specifications		
3468	2015		47270	Cardiac resynchronization therapy implantable defibrillator	CT2866	Implantable cardiac electrical stimulators			A26	Insufficient Device Problem Information		
3469	2015	All other reportable incidents			CT257	Implantable cardiac pacemakers			A07	Electrical /Electronic Property Problem		
									A0709	Device Sensing Problem		
3470	2016	Unanticipated serious deterioration in state of health	35852	Single-chamber implantable defibrillator	CT2866	Implantable cardiac electrical stimulators			A07	Electrical /Electronic Property Problem		
									A0709	Device Sensing Problem		
3471	2015				CT1684	Cardiac pacemakers and associated devices			A09	Output Problem		
									A0907	No Device Output		
3472	2016	TBD	35223	Endocardial/interventricular septal pacing lead	CT1684	Cardiac pacemakers and associated devices			A26	Insufficient Device Problem Information		
3473	2016	TBD			CT257	Implantable cardiac pacemakers			A07	Electrical /Electronic Property Problem		

									A0708	Power Problem		
3474	2016	TBD			CT257	Implantable cardiac pacemakers			A07	Electrical /Electronic Property Problem		
									A0709	Device Sensing Problem		
									A09	Output Problem		
									A090404	Output Above Specifications		
3475	2015				CT257	Implantable cardiac pacemakers			A07	Electrical /Electronic Property Problem		
									A0708	Power Problem		
3476	2015	All other reportable incidents	47270	Cardiac resynchronization therapy implantable defibrillator	CT2866	Implantable cardiac electrical stimulators			A07	Electrical /Electronic Property Problem	23/10/2015	
									A0708	Power Problem	23/10/2015	
3477	2016	All other reportable incidents	47263	Cardiac resynchronization therapy implantable pacemaker	CT1684	Cardiac pacemakers and associated devices			A07	Electrical /Electronic Property Problem	08/01/2016	
									A0708	Power Problem	08/01/2016	
					CT1685	Cardiac pacemakers			A07	Electrical /Electronic Property Problem	08/01/2016	
									A0708	Power Problem	08/01/2016	
					CT257	Implantable cardiac pacemakers			A07	Electrical /Electronic Property Problem	08/01/2016	
									A0708	Power Problem	08/01/2016	
					CT2866	Implantable cardiac electrical stimulators			A07	Electrical /Electronic Property Problem	08/01/2016	
									A0708	Power Problem	08/01/2016	
3478	2015	All other reportable incidents	47270	Cardiac resynchronization therapy implantable defibrillator	CT2866	Implantable cardiac electrical stimulators			A07	Electrical /Electronic Property Problem		
									A0708	Power Problem		
3479	2015				CT257	Implantable cardiac pacemakers			A26	Insufficient Device Problem Information		
3480	2015		47270	Cardiac resynchronization therapy implantable defibrillator	CT2866	Implantable cardiac electrical stimulators			A07	Electrical /Electronic Property Problem		
									A0708	Power Problem		
3481	2016	TBD			CT257	Implantable cardiac pacemakers			A07	Electrical /Electronic Property Problem		
									A0708	Power Problem		
3482	2016	All other reportable incidents	35223	Endocardial/interventricular septal pacing lead	CT1684	Cardiac pacemakers and associated devices			A07	Electrical /Electronic Property Problem		
									A0709	Device Sensing Problem		
3483	2016	All other reportable incidents	47263	Cardiac resynchronization therapy implantable pacemaker	CT1684	Cardiac pacemakers and associated devices			A07	Electrical /Electronic Property Problem	25/05/2016	
									A0708	Power Problem	25/05/2016	
					CT1685	Cardiac pacemakers			A07	Electrical /Electronic Property Problem	25/05/2016	
									A0708	Power Problem	25/05/2016	
					CT257	Implantable cardiac pacemakers			A07	Electrical /Electronic Property Problem	25/05/2016	
									A0708	Power Problem	25/05/2016	
					CT2866	Implantable cardiac electrical stimulators			A07	Electrical /Electronic Property Problem	25/05/2016	
									A0708	Power Problem	25/05/2016	
3484	2015		47270	Cardiac resynchronization therapy implantable defibrillator	CT2866	Implantable cardiac electrical stimulators			A07	Electrical /Electronic Property Problem		
									A0708	Power Problem		
3485	2015		47270	Cardiac resynchronization therapy implantable defibrillator	CT2866	Implantable cardiac electrical stimulators			A26	Insufficient Device Problem Information		
3486	2015				CT257	Implantable cardiac pacemakers			A26	Insufficient Device Problem Information		
3487	2015				CT257	Implantable cardiac pacemakers			A07	Electrical /Electronic Property Problem		
									A0708	Power Problem		
3488	2016	TBD	47270	Cardiac resynchronization therapy implantable defibrillator	CT2866	Implantable cardiac electrical stimulators			A07	Electrical /Electronic Property Problem		
									A0709	Device Sensing Problem		
3489	2015	All other reportable incidents	47270	Cardiac resynchronization therapy implantable defibrillator	CT2866	Implantable cardiac electrical stimulators			A07	Electrical /Electronic Property Problem	17/06/2015	
									A0708	Power Problem	17/06/2015	
3490	2016	All other reportable incidents			CT257	Implantable cardiac pacemakers			A07	Electrical /Electronic Property Problem	10/03/2016	
									A0708	Power Problem	10/03/2016	
3491	2015		47270	Cardiac resynchronization therapy implantable defibrillator	CT2866	Implantable cardiac electrical stimulators			A13	Communication or Transmission Problem		
									A26	Insufficient Device Problem Information		
3492	2015		47270	Cardiac resynchronization therapy implantable defibrillator	CT2866	Implantable cardiac electrical stimulators			A07	Electrical /Electronic Property Problem		
									A0708	Power Problem		
3493	2016	All other reportable incidents			CT257	Implantable cardiac pacemakers			A07	Electrical /Electronic Property Problem		
									A0708	Power Problem		
									A13	Communication or Transmission Problem		
									A26	Insufficient Device Problem Information		
3494	2015				CT257	Implantable cardiac pacemakers			A07	Electrical /Electronic Property Problem		
									A0708	Power Problem		
3495	2015	Death	47270	Cardiac resynchronization therapy implantable defibrillator	CT2866	Implantable cardiac electrical stimulators			A24	No Apparent Device Problem	20/12/2014	
									A26	Insufficient Device Problem Information	20/12/2014	
3496	2015		47270	Cardiac resynchronization therapy implantable defibrillator	CT2866	Implantable cardiac electrical stimulators			A26	Insufficient Device Problem Information		
3497	2016	TBD	35223	Endocardial/interventricular septal pacing lead	CT1684	Cardiac pacemakers and associated devices			A26	Insufficient Device Problem Information		
3498	2015		47270	Cardiac resynchronization therapy implantable defibrillator	CT2866	Implantable cardiac electrical stimulators			A07	Electrical /Electronic Property Problem		
									A0708	Power Problem		
3499	2015	All other reportable incidents			CT257	Implantable cardiac pacemakers			A07	Electrical /Electronic Property Problem	28/05/2015	
									A0709	Device Sensing Problem	28/05/2015	
3500	2016	TBD			CT257	Implantable cardiac pacemakers			A17	Compatibility Problem		
									A1702	Device-Device Incompatibility		
3501	2015				CT257	Implantable cardiac pacemakers			A07	Electrical /Electronic Property Problem		
									A0708	Power Problem		
3502	2015				CT257	Implantable cardiac pacemakers			A07	Electrical /Electronic Property Problem		
									A0708	Power Problem		
3503	2015		47270	Cardiac resynchronization therapy implantable defibrillator	CT2866	Implantable cardiac electrical stimulators			A07	Electrical /Electronic Property Problem		
									A0708	Power Problem		
3504	2016	All other reportable incidents	47263	Cardiac resynchronization therapy implantable pacemaker	CT1684	Cardiac pacemakers and associated devices			A07	Electrical /Electronic Property Problem	20/01/2016	
									A0708	Power Problem	20/01/2016	
					CT1685	Cardiac pacemakers			A07	Electrical /Electronic Property Problem	20/01/2016	
									A0708	Power Problem	20/01/2016	
					CT257	Implantable cardiac pacemakers			A07	Electrical /Electronic Property Problem	20/01/2016	
									A0708	Power Problem	20/01/2016	
					CT2866	Implantable cardiac electrical stimulators			A07	Electrical /Electronic Property Problem	20/01/2016	
									A0708	Power Problem	20/01/2016	
3505	2015				CT257	Implantable cardiac pacemakers			A26	Insufficient Device Problem Information		
3506	2015	All other reportable incidents	47270	Cardiac resynchronization therapy implantable defibrillator	CT2866	Implantable cardiac electrical stimulators			A07	Electrical /Electronic Property Problem	13/10/2015	
									A0708	Power Problem	13/10/2015	
3507	2015		35852	Single-chamber implantable defibrillator	CT2866	Implantable cardiac electrical stimulators			A07	Electrical /Electronic Property Problem		
									A0708	Power Problem		

3508	2015		47270	Cardiac resynchronization therapy implantable defibrillator	CT2866	Implantable cardiac electrical stimulators			A11	Computer Software Problem		
									A1102	Application Program Problem		
3509	2016	All other reportable incidents	47270	Cardiac resynchronization therapy implantable defibrillator	CT2866	Implantable cardiac electrical stimulators			A07	Electrical /Electronic Property Problem	17/08/2016	
									A0708	Power Problem	17/08/2016	
3510	2016	All other reportable incidents	47270	Cardiac resynchronization therapy implantable defibrillator	CT2866	Implantable cardiac electrical stimulators			A07	Electrical /Electronic Property Problem	19/08/2016	
									A0708	Power Problem	19/08/2016	
3511	2015				CT257	Implantable cardiac pacemakers			A26	Insufficient Device Problem Information		
3512	2015				CT1684	Cardiac pacemakers and associated devices			A15	Activation, Positioning or Separation Problem		
									A150101	Activation Failure		
3513	2015				CT257	Implantable cardiac pacemakers			A07	Electrical /Electronic Property Problem		
									A0709	Device Sensing Problem		
3514	2015				CT1684	Cardiac pacemakers and associated devices			A07	Electrical /Electronic Property Problem		
									A0708	Power Problem		
									A15	Activation, Positioning or Separation Problem		
									A150101	Activation Failure		
3515	2015				CT257	Implantable cardiac pacemakers			A07	Electrical /Electronic Property Problem	28/05/2015	
									A0708	Power Problem	28/05/2015	
3516	2016	All other reportable incidents	47263	Cardiac resynchronization therapy implantable pacemaker	CT1684	Cardiac pacemakers and associated devices			A07	Electrical /Electronic Property Problem	14/01/2016	
									A0708	Power Problem	14/01/2016	
					CT1685	Cardiac pacemakers			A07	Electrical /Electronic Property Problem	14/01/2016	
									A0708	Power Problem	14/01/2016	
					CT257	Implantable cardiac pacemakers			A07	Electrical /Electronic Property Problem	14/01/2016	
									A0708	Power Problem	14/01/2016	
					CT2866	Implantable cardiac electrical stimulators			A07	Electrical /Electronic Property Problem	14/01/2016	
									A0708	Power Problem	14/01/2016	
3517	2015				CT257	Implantable cardiac pacemakers			A26	Insufficient Device Problem Information		
3518	2016	All other reportable incidents	47263	Cardiac resynchronization therapy implantable pacemaker	CT1684	Cardiac pacemakers and associated devices			A07	Electrical /Electronic Property Problem	11/12/2015	
									A0708	Power Problem	11/12/2015	
									A09	Output Problem	11/12/2015	
									A0907	No Device Output	11/12/2015	
					CT1685	Cardiac pacemakers			A07	Electrical /Electronic Property Problem	11/12/2015	
									A0708	Power Problem	11/12/2015	
									A09	Output Problem	11/12/2015	
									A0907	No Device Output	11/12/2015	
					CT257	Implantable cardiac pacemakers			A07	Electrical /Electronic Property Problem	11/12/2015	
									A0708	Power Problem	11/12/2015	
									A09	Output Problem	11/12/2015	
									A0907	No Device Output	11/12/2015	
					CT2866	Implantable cardiac electrical stimulators			A07	Electrical /Electronic Property Problem	11/12/2015	
									A0708	Power Problem	11/12/2015	
									A09	Output Problem	11/12/2015	
									A0907	No Device Output	11/12/2015	
3519	2016	All other reportable incidents	47263	Cardiac resynchronization therapy implantable pacemaker	CT1684	Cardiac pacemakers and associated devices			A07	Electrical /Electronic Property Problem	08/01/2016	
									A0708	Power Problem	08/01/2016	
					CT1685	Cardiac pacemakers			A07	Electrical /Electronic Property Problem	08/01/2016	
									A0708	Power Problem	08/01/2016	
					CT257	Implantable cardiac pacemakers			A07	Electrical /Electronic Property Problem	08/01/2016	
									A0708	Power Problem	08/01/2016	
					CT2866	Implantable cardiac electrical stimulators			A07	Electrical /Electronic Property Problem	08/01/2016	
									A0708	Power Problem	08/01/2016	
3520	2015	All other reportable incidents	47270	Cardiac resynchronization therapy implantable defibrillator	CT2866	Implantable cardiac electrical stimulators			A07	Electrical /Electronic Property Problem	01/09/2015	
									A0708	Power Problem	01/09/2015	
3521	2015		47270	Cardiac resynchronization therapy implantable defibrillator	CT2866	Implantable cardiac electrical stimulators			A26	Insufficient Device Problem Information		
3522	2016	TBD	35223	Endocardial/interventricular septal pacing lead	CT1684	Cardiac pacemakers and associated devices			A15	Activation, Positioning or Separation Problem		
									A150203	Difficult or Delayed Positioning		
3523	2015				CT257	Implantable cardiac pacemakers			A07	Electrical /Electronic Property Problem		
									A0708	Power Problem		
3524	2016	TBD			CT257	Implantable cardiac pacemakers			A07	Electrical /Electronic Property Problem		
									A0708	Power Problem		
3525	2015				CT257	Implantable cardiac pacemakers			A07	Electrical /Electronic Property Problem		
									A0709	Device Sensing Problem		
3526	2015				CT257	Implantable cardiac pacemakers			A07	Electrical /Electronic Property Problem	23/06/2015	
									A0708	Power Problem	23/06/2015	
3527	2015		47270	Cardiac resynchronization therapy implantable defibrillator	CT2866	Implantable cardiac electrical stimulators			A07	Electrical /Electronic Property Problem		
									A0708	Power Problem		
3528	2016	TBD	47270	Cardiac resynchronization therapy implantable defibrillator	CT2866	Implantable cardiac electrical stimulators			A07	Electrical /Electronic Property Problem		
									A0708	Power Problem		
3529	2015	All other reportable incidents	36104	Implantable pacemaker analyser	CT1684	Cardiac pacemakers and associated devices			A10	Temperature Problem		
									A1005	Overheating of Device		
3530	2015		47270	Cardiac resynchronization therapy implantable defibrillator	CT2866	Implantable cardiac electrical stimulators			A07	Electrical /Electronic Property Problem		
									A0708	Power Problem		
3531	2015				CT257	Implantable cardiac pacemakers			A07	Electrical /Electronic Property Problem		
									A0708	Power Problem		
3532	2015	All other reportable incidents			CT257	Implantable cardiac pacemakers			A07	Electrical /Electronic Property Problem	22/06/2015	
									A0708	Power Problem	22/06/2015	
3533	2015	All other reportable incidents	47270	Cardiac resynchronization therapy implantable defibrillator	CT2866	Implantable cardiac electrical stimulators			A13	Communication or Transmission Problem	29/05/2015	
									A26	Insufficient Device Problem Information	29/05/2015	
3534	2016	TBD	35223	Endocardial/interventricular septal pacing lead	CT1684	Cardiac pacemakers and associated devices			A26	Insufficient Device Problem Information		
3535	2016	TBD			CT257	Implantable cardiac pacemakers			A07	Electrical /Electronic Property Problem		
									A0708	Power Problem		
3536	2015		47270	Cardiac resynchronization therapy implantable defibrillator	CT2866	Implantable cardiac electrical stimulators			A26	Insufficient Device Problem Information		
3537	2015				CT1684	Cardiac pacemakers and associated devices			A15	Activation, Positioning or Separation Problem		
									A150101	Activation Failure		
3538	2015				CT1684	Cardiac pacemakers and associated devices			A15	Activation, Positioning or Separation Problem		
									A150101	Activation Failure		

3539	2016	All other reportable incidents	35852	Single-chamber implantable defibrillator	CT2866	Implantable cardiac electrical stimulators			A07	Electrical /Electronic Property Problem		
									A0708	Power Problem		
3540	2015				CT257	Implantable cardiac pacemakers			A07	Electrical /Electronic Property Problem		
									A0709	Device Sensing Problem		
3541	2015		47270	Cardiac resynchronization therapy implantable defibrillator	CT2866	Implantable cardiac electrical stimulators			A26	Insufficient Device Problem Information		
3542	2015		47270	Cardiac resynchronization therapy implantable defibrillator	CT2866	Implantable cardiac electrical stimulators			A26	Insufficient Device Problem Information		
3543	2016	TBD			CT1684	Cardiac pacemakers and associated devices			A15	Activation, Positioning or Separation Problem		
									A150101	Activation Failure		
3544	2016	TBD	47270	Cardiac resynchronization therapy implantable defibrillator	CT2866	Implantable cardiac electrical stimulators			A07	Electrical /Electronic Property Problem		
									A0708	Power Problem		
3545	2016	TBD	47270	Cardiac resynchronization therapy implantable defibrillator	CT2866	Implantable cardiac electrical stimulators			A07	Electrical /Electronic Property Problem		
									A0708	Power Problem		
3546	2015				CT257	Implantable cardiac pacemakers			A26	Insufficient Device Problem Information		
3547	2015	All other reportable incidents			CT257	Implantable cardiac pacemakers			A07	Electrical /Electronic Property Problem		
									A0708	Power Problem		
3548	2015	All other reportable incidents			CT1684	Cardiac pacemakers and associated devices			A09	Output Problem		
									A0908	Incorrect, Inadequate or Imprecise Result or Readings		
3549	2015		47270	Cardiac resynchronization therapy implantable defibrillator	CT2866	Implantable cardiac electrical stimulators			A26	Insufficient Device Problem Information		
3550	2016	All other reportable incidents	47270	Cardiac resynchronization therapy implantable defibrillator	CT2866	Implantable cardiac electrical stimulators			A07	Electrical /Electronic Property Problem	14/04/2016	
									A0703	Arcing	14/04/2016	
3551	2015				CT257	Implantable cardiac pacemakers			A07	Electrical /Electronic Property Problem		
									A0708	Power Problem		
3552	2015				CT1684	Cardiac pacemakers and associated devices			A1104	Computer Operating System Problem		
					CT257	Implantable cardiac pacemakers			A1104	Computer Operating System Problem		
3553	2015		47270	Cardiac resynchronization therapy implantable defibrillator	CT2866	Implantable cardiac electrical stimulators			A26	Insufficient Device Problem Information		
3554	2015		47270	Cardiac resynchronization therapy implantable defibrillator	CT2866	Implantable cardiac electrical stimulators			A26	Insufficient Device Problem Information		
3555	2015				CT1684	Cardiac pacemakers and associated devices			A09	Output Problem		
									A0907	No Device Output		
3556	2015	All other reportable incidents			CT1684	Cardiac pacemakers and associated devices			A07	Electrical /Electronic Property Problem		
									A0708	Power Problem		
3557	2016	TBD	35852	Single-chamber implantable defibrillator	CT2866	Implantable cardiac electrical stimulators			A07	Electrical /Electronic Property Problem		
									A0709	Device Sensing Problem		
3558	2015				CT1684	Cardiac pacemakers and associated devices			A09	Output Problem		
									A0907	No Device Output		
3559	2016	All other reportable incidents			CT257	Implantable cardiac pacemakers			A07	Electrical /Electronic Property Problem		
									A0708	Power Problem		
3560	2016	TBD	35223	Endocardial/interventricular septal pacing lead	CT1684	Cardiac pacemakers and associated devices			A26	Insufficient Device Problem Information		
3561	2015		47270	Cardiac resynchronization therapy implantable defibrillator	CT2866	Implantable cardiac electrical stimulators			A11	Computer Software Problem		
									A1102	Application Program Problem		
3562	2016	All other reportable incidents	47270	Cardiac resynchronization therapy implantable defibrillator	CT2866	Implantable cardiac electrical stimulators			A07	Electrical /Electronic Property Problem	31/08/2016	
									A0708	Power Problem	31/08/2016	
3563	2015	Death			CT257	Implantable cardiac pacemakers			A09	Output Problem		
									A0907	No Device Output		
3564	2015				CT257	Implantable cardiac pacemakers			A26	Insufficient Device Problem Information		
3565	2015				CT257	Implantable cardiac pacemakers			A26	Insufficient Device Problem Information		
3566	2015				CT257	Implantable cardiac pacemakers			A07	Electrical /Electronic Property Problem		
									A0708	Power Problem		
3567	2015	All other reportable incidents			CT257	Implantable cardiac pacemakers			A07	Electrical /Electronic Property Problem		
									A0708	Power Problem		
3568	2016	All other reportable incidents			CT257	Implantable cardiac pacemakers			A07	Electrical /Electronic Property Problem	20/04/2016	
									A0708	Power Problem	20/04/2016	
3569	2016	Death	47270	Cardiac resynchronization therapy implantable defibrillator	CT2866	Implantable cardiac electrical stimulators			A26	Insufficient Device Problem Information		
3570	2016	TBD	35223	Endocardial/interventricular septal pacing lead	CT1684	Cardiac pacemakers and associated devices			A26	Insufficient Device Problem Information		
3571	2015				CT257	Implantable cardiac pacemakers			A26	Insufficient Device Problem Information		
3572	2015				CT257	Implantable cardiac pacemakers			A26	Insufficient Device Problem Information		
3573	2015	All other reportable incidents	47263	Cardiac resynchronization therapy implantable pacemaker	CT1684	Cardiac pacemakers and associated devices			A07	Electrical /Electronic Property Problem	12/10/2015	
									A0708	Power Problem	12/10/2015	
					CT1685	Cardiac pacemakers			A07	Electrical /Electronic Property Problem	12/10/2015	
									A0708	Power Problem	12/10/2015	
					CT257	Implantable cardiac pacemakers			A07	Electrical /Electronic Property Problem	12/10/2015	
									A0708	Power Problem	12/10/2015	
					CT2866	Implantable cardiac electrical stimulators			A07	Electrical /Electronic Property Problem	12/10/2015	
									A0708	Power Problem	12/10/2015	
3574	2015	All other reportable incidents	47270	Cardiac resynchronization therapy implantable defibrillator	CT2866	Implantable cardiac electrical stimulators			A07	Electrical /Electronic Property Problem		
									A0709	Device Sensing Problem		
3575	2015		47270	Cardiac resynchronization therapy implantable defibrillator	CT2866	Implantable cardiac electrical stimulators			A26	Insufficient Device Problem Information		
3576	2015				CT257	Implantable cardiac pacemakers			A07	Electrical /Electronic Property Problem	01/06/2015	
									A0708	Power Problem	01/06/2015	
3577	2015		47270	Cardiac resynchronization therapy implantable defibrillator	CT2866	Implantable cardiac electrical stimulators			A07	Electrical /Electronic Property Problem		
									A0708	Power Problem		
									A0709	Device Sensing Problem		
3578	2015		47270	Cardiac resynchronization therapy implantable defibrillator	CT2866	Implantable cardiac electrical stimulators			A26	Insufficient Device Problem Information		
3579	2015	All other reportable incidents	47270	Cardiac resynchronization therapy implantable defibrillator	CT2866	Implantable cardiac electrical stimulators			A07	Electrical /Electronic Property Problem	04/11/2015	
									A0708	Power Problem	04/11/2015	
3580	2015		35852	Single-chamber implantable defibrillator	CT2866	Implantable cardiac electrical stimulators			A07	Electrical /Electronic Property Problem		
									A0709	Device Sensing Problem		
3581	2015		47270	Cardiac resynchronization therapy implantable defibrillator	CT2866	Implantable cardiac electrical stimulators			A26	Insufficient Device Problem Information		
3582	2015		35223	Endocardial/interventricular septal pacing lead	CT1684	Cardiac pacemakers and associated devices			A07	Electrical /Electronic Property Problem		
									A0709	Device Sensing Problem		

3583	2016	All other reportable incidents	47270	Cardiac resynchronization therapy implantable defibrillator	CT2866	Implantable cardiac electrical stimulators			A07	Electrical /Electronic Property Problem	19/08/2016	
									A0709	Device Sensing Problem	19/08/2016	
3584	2015				CT257	Implantable cardiac pacemakers			A07	Electrical /Electronic Property Problem		
									A0709	Device Sensing Problem		
3585	2016	TBD	47270	Cardiac resynchronization therapy implantable defibrillator	CT2866	Implantable cardiac electrical stimulators			A07	Electrical /Electronic Property Problem		
									A0708	Power Problem		
3586	2015				CT257	Implantable cardiac pacemakers			A26	Insufficient Device Problem Information		
3587	2016	TBD	35852	Single-chamber implantable defibrillator	CT2866	Implantable cardiac electrical stimulators			A07	Electrical /Electronic Property Problem		
									A0709	Device Sensing Problem		
3588	2015		47270	Cardiac resynchronization therapy implantable defibrillator	CT2866	Implantable cardiac electrical stimulators			A26	Insufficient Device Problem Information		
3589	2015		47270	Cardiac resynchronization therapy implantable defibrillator	CT2866	Implantable cardiac electrical stimulators			A07	Electrical /Electronic Property Problem		
									A0709	Device Sensing Problem		
3590	2016	TBD			CT257	Implantable cardiac pacemakers			A07	Electrical /Electronic Property Problem		
									A0709	Device Sensing Problem		
3591	2016	TBD			CT257	Implantable cardiac pacemakers			A07	Electrical /Electronic Property Problem		
									A0708	Power Problem		
3592	2016	All other reportable incidents	47270	Cardiac resynchronization therapy implantable defibrillator	CT2866	Implantable cardiac electrical stimulators			A07	Electrical /Electronic Property Problem		
									A0708	Power Problem		
3593	2016	All other reportable incidents	47270	Cardiac resynchronization therapy implantable defibrillator	CT2866	Implantable cardiac electrical stimulators			A07	Electrical /Electronic Property Problem		
									A0703	Arcing		
3594	2016	TBD	47270	Cardiac resynchronization therapy implantable defibrillator	CT2866	Implantable cardiac electrical stimulators			A07	Electrical /Electronic Property Problem		
									A0709	Device Sensing Problem		
3595	2016	TBD	35223	Endocardial/interventricular septal pacing lead	CT1684	Cardiac pacemakers and associated devices			A26	Insufficient Device Problem Information		
3596	2015		47270	Cardiac resynchronization therapy implantable defibrillator	CT2866	Implantable cardiac electrical stimulators			A07	Electrical /Electronic Property Problem		
									A0708	Power Problem		
3597	2015				CT257	Implantable cardiac pacemakers			A26	Insufficient Device Problem Information		
3598	2016	TBD	47270	Cardiac resynchronization therapy implantable defibrillator	CT2866	Implantable cardiac electrical stimulators			A07	Electrical /Electronic Property Problem		
									A0708	Power Problem		
3599	2015		47270	Cardiac resynchronization therapy implantable defibrillator	CT2866	Implantable cardiac electrical stimulators			A26	Insufficient Device Problem Information		
3600	2015	All other reportable incidents			CT257	Implantable cardiac pacemakers			A07	Electrical /Electronic Property Problem	27/11/2015	
									A0708	Power Problem	27/11/2015	
3601	2015				CT257	Implantable cardiac pacemakers			A26	Insufficient Device Problem Information		
3602	2015	All other reportable incidents			CT257	Implantable cardiac pacemakers			A07	Electrical /Electronic Property Problem	12/11/2015	
									A0708	Power Problem	12/11/2015	
3603	2015		47270	Cardiac resynchronization therapy implantable defibrillator	CT2866	Implantable cardiac electrical stimulators			A1304	Telemetry Discrepancy		
									A26	Insufficient Device Problem Information		
3604	2015		47270	Cardiac resynchronization therapy implantable defibrillator	CT2866	Implantable cardiac electrical stimulators			A26	Insufficient Device Problem Information		
3605	2015	All other reportable incidents	47263	Cardiac resynchronization therapy implantable pacemaker	CT1684	Cardiac pacemakers and associated devices			A07	Electrical /Electronic Property Problem		
									A0708	Power Problem		
					CT1685	Cardiac pacemakers			A07	Electrical /Electronic Property Problem		
									A0708	Power Problem		
					CT257	Implantable cardiac pacemakers			A07	Electrical /Electronic Property Problem		
									A0708	Power Problem		
					CT2866	Implantable cardiac electrical stimulators			A07	Electrical /Electronic Property Problem		
									A0708	Power Problem		
3606	2016	TBD			CT257	Implantable cardiac pacemakers			A07	Electrical /Electronic Property Problem		
									A0708	Power Problem		
3607	2015				CT257	Implantable cardiac pacemakers			A26	Insufficient Device Problem Information		
3608	2015				CT257	Implantable cardiac pacemakers			A07	Electrical /Electronic Property Problem		
									A0708	Power Problem		
3609	2015	All other reportable incidents	47270	Cardiac resynchronization therapy implantable defibrillator	CT2866	Implantable cardiac electrical stimulators			A07	Electrical /Electronic Property Problem		
									A0708	Power Problem		
3610	2015				CT257	Implantable cardiac pacemakers			A07	Electrical /Electronic Property Problem	13/05/2014	
									A0709	Device Sensing Problem	13/05/2014	
3611	2015	All other reportable incidents	47263	Cardiac resynchronization therapy implantable pacemaker	CT1684	Cardiac pacemakers and associated devices			A07	Electrical /Electronic Property Problem	15/06/2015	
									A0709	Device Sensing Problem	15/06/2015	
					CT1685	Cardiac pacemakers			A07	Electrical /Electronic Property Problem	15/06/2015	
									A0709	Device Sensing Problem	15/06/2015	
					CT257	Implantable cardiac pacemakers			A07	Electrical /Electronic Property Problem	15/06/2015	
									A0709	Device Sensing Problem	15/06/2015	
					CT2866	Implantable cardiac electrical stimulators			A07	Electrical /Electronic Property Problem	15/06/2015	
									A0709	Device Sensing Problem	15/06/2015	
3612	2016	All other reportable incidents	47270	Cardiac resynchronization therapy implantable defibrillator	CT2866	Implantable cardiac electrical stimulators			A07	Electrical /Electronic Property Problem	15/04/2016	
									A0703	Arcing	15/04/2016	
3613	2015				CT257	Implantable cardiac pacemakers			A07	Electrical /Electronic Property Problem		
									A0708	Power Problem		
3614	2016	TBD			CT257	Implantable cardiac pacemakers			A07	Electrical /Electronic Property Problem		
									A0709	Device Sensing Problem		
3615	2016	All other reportable incidents			CT1684	Cardiac pacemakers and associated devices			A09	Output Problem		
									A0907	No Device Output		
3616	2015		47270	Cardiac resynchronization therapy implantable defibrillator	CT2866	Implantable cardiac electrical stimulators			A26	Insufficient Device Problem Information		
3617	2016	TBD	47270	Cardiac resynchronization therapy implantable defibrillator	CT2866	Implantable cardiac electrical stimulators			A07	Electrical /Electronic Property Problem		
									A0708	Power Problem		
3618	2016	TBD	47270	Cardiac resynchronization therapy implantable defibrillator	CT2866	Implantable cardiac electrical stimulators			A07	Electrical /Electronic Property Problem		
									A0708	Power Problem		
3619	2015	All other reportable incidents			CT257	Implantable cardiac pacemakers			A07	Electrical /Electronic Property Problem		
									A0708	Power Problem		
3620	2016	TBD	47270	Cardiac resynchronization therapy implantable defibrillator	CT2866	Implantable cardiac electrical stimulators			A07	Electrical /Electronic Property Problem		
									A0709	Device Sensing Problem		
3621	2016	TBD	35223	Endocardial/interventricular septal pacing lead	CT1684	Cardiac pacemakers and associated devices			A26	Insufficient Device Problem Information		
3622	2016	All other reportable incidents	47270	Cardiac resynchronization therapy implantable defibrillator	CT2866	Implantable cardiac electrical stimulators			A07	Electrical /Electronic Property Problem	17/02/2016	
									A0708	Power Problem	17/02/2016	

3623	2015		47270	Cardiac resynchronization therapy implantable defibrillator	CT2866	Implantable cardiac electrical stimulators			A26	Insufficient Device Problem Information		
3624	2015		47270	Cardiac resynchronization therapy implantable defibrillator	CT2866	Implantable cardiac electrical stimulators			A26	Insufficient Device Problem Information		
3625	2015		47270	Cardiac resynchronization therapy implantable defibrillator	CT2866	Implantable cardiac electrical stimulators			A07	Electrical /Electronic Property Problem		
									A0709	Device Sensing Problem		
3626	2015	All other reportable incidents	47270	Cardiac resynchronization therapy implantable defibrillator	CT2866	Implantable cardiac electrical stimulators			A07	Electrical /Electronic Property Problem		
									A0708	Power Problem		
3627	2015	All other reportable incidents	47263	Cardiac resynchronization therapy implantable pacemaker	CT1684	Cardiac pacemakers and associated devices			A13	Communication or Transmission Problem		
									A26	Insufficient Device Problem Information		
					CT1685	Cardiac pacemakers			A13	Communication or Transmission Problem		
									A26	Insufficient Device Problem Information		
					CT257	Implantable cardiac pacemakers			A13	Communication or Transmission Problem		
									A26	Insufficient Device Problem Information		
					CT2866	Implantable cardiac electrical stimulators			A13	Communication or Transmission Problem		
									A26	Insufficient Device Problem Information		
3628	2016	TBD	47270	Cardiac resynchronization therapy implantable defibrillator	CT2866	Implantable cardiac electrical stimulators			A09	Output Problem		
									A090404	Output Above Specifications		
3629	2016	All other reportable incidents			CT257	Implantable cardiac pacemakers			A07	Electrical /Electronic Property Problem		
									A0709	Device Sensing Problem		
3630	2015		47270	Cardiac resynchronization therapy implantable defibrillator	CT2866	Implantable cardiac electrical stimulators			A07	Electrical /Electronic Property Problem		
									A0708	Power Problem		
3631	2016	TBD	35223	Endocardial/interventricular septal pacing lead	CT1684	Cardiac pacemakers and associated devices			A26	Insufficient Device Problem Information		
3632	2016	TBD	35223	Endocardial/interventricular septal pacing lead	CT1684	Cardiac pacemakers and associated devices			A26	Insufficient Device Problem Information		
3633	2016	TBD			CT1684	Cardiac pacemakers and associated devices			A09	Output Problem		
									A0908	Incorrect, Inadequate or Imprecise Result or Readings		
3634	2015				CT257	Implantable cardiac pacemakers			A26	Insufficient Device Problem Information		
3635	2016	TBD	35223	Endocardial/interventricular septal pacing lead	CT1684	Cardiac pacemakers and associated devices			A07	Electrical /Electronic Property Problem		
									A0709	Device Sensing Problem		
3636	2016	All other reportable incidents			CT257	Implantable cardiac pacemakers			A07	Electrical /Electronic Property Problem	14/09/2016	
									A0708	Power Problem	14/09/2016	
3637	2016	TBD	47270	Cardiac resynchronization therapy implantable defibrillator	CT2866	Implantable cardiac electrical stimulators			A07	Electrical /Electronic Property Problem		
									A0709	Device Sensing Problem		
3638	2016	TBD			CT257	Implantable cardiac pacemakers			A07	Electrical /Electronic Property Problem		
									A0709	Device Sensing Problem		
3639	2015				CT257	Implantable cardiac pacemakers			A07	Electrical /Electronic Property Problem		
									A0708	Power Problem		
3640	2015		47270	Cardiac resynchronization therapy implantable defibrillator	CT2866	Implantable cardiac electrical stimulators			A07	Electrical /Electronic Property Problem		
									A0708	Power Problem		
3641	2015		47270	Cardiac resynchronization therapy implantable defibrillator	CT2866	Implantable cardiac electrical stimulators			A26	Insufficient Device Problem Information		
3642	2015		47270	Cardiac resynchronization therapy implantable defibrillator	CT2866	Implantable cardiac electrical stimulators			A26	Insufficient Device Problem Information		
3643	2015	All other reportable incidents	47270	Cardiac resynchronization therapy implantable defibrillator	CT2866	Implantable cardiac electrical stimulators			A07	Electrical /Electronic Property Problem		
									A0708	Power Problem		
3644	2015		35852	Single-chamber implantable defibrillator	CT2866	Implantable cardiac electrical stimulators			A01	Patient Device Interaction Problem		
									A17	Compatibility Problem		
3645	2015				CT257	Implantable cardiac pacemakers			A07	Electrical /Electronic Property Problem		
									A0708	Power Problem		
3646	2015	All other reportable incidents	47263	Cardiac resynchronization therapy implantable pacemaker	CT1684	Cardiac pacemakers and associated devices			A07	Electrical /Electronic Property Problem		
									A0708	Power Problem		
					CT1685	Cardiac pacemakers			A07	Electrical /Electronic Property Problem		
									A0708	Power Problem		
					CT257	Implantable cardiac pacemakers			A07	Electrical /Electronic Property Problem		
									A0708	Power Problem		
					CT2866	Implantable cardiac electrical stimulators			A07	Electrical /Electronic Property Problem		
									A0708	Power Problem		
3647	2015	All other reportable incidents			CT257	Implantable cardiac pacemakers			A11	Computer Software Problem		
									A1102	Application Program Problem		
									A13	Communication or Transmission Problem		
									A26	Insufficient Device Problem Information		
3648	2015		35223	Endocardial/interventricular septal pacing lead	CT1684	Cardiac pacemakers and associated devices			A0104	Migration or Expulsion of Device		
3649	2016	TBD	35223	Endocardial/interventricular septal pacing lead	CT1684	Cardiac pacemakers and associated devices			A26	Insufficient Device Problem Information		
3650	2016	All other reportable incidents	47270	Cardiac resynchronization therapy implantable defibrillator	CT2866	Implantable cardiac electrical stimulators			A07	Electrical /Electronic Property Problem		
									A0708	Power Problem		
3651	2015	All other reportable incidents	35223	Endocardial/interventricular septal pacing lead	CT1684	Cardiac pacemakers and associated devices			A07	Electrical /Electronic Property Problem		
									A0709	Device Sensing Problem		
3652	2016	TBD	47270	Cardiac resynchronization therapy implantable defibrillator	CT2866	Implantable cardiac electrical stimulators			A11	Computer Software Problem		
									A1102	Application Program Problem		
3653	2015		47270	Cardiac resynchronization therapy implantable defibrillator	CT2866	Implantable cardiac electrical stimulators			A07	Electrical /Electronic Property Problem		
									A0708	Power Problem		
3654	2016	TBD	47270	Cardiac resynchronization therapy implantable defibrillator	CT2866	Implantable cardiac electrical stimulators			A07	Electrical /Electronic Property Problem		
									A0708	Power Problem		
3655	2015				CT257	Implantable cardiac pacemakers			A26	Insufficient Device Problem Information		
3656	2015	All other reportable incidents			CT257	Implantable cardiac pacemakers			A07	Electrical /Electronic Property Problem	03/12/2015	
									A0708	Power Problem	03/12/2015	
3657	2016	All other reportable incidents			CT257	Implantable cardiac pacemakers			A07	Electrical /Electronic Property Problem		
									A0708	Power Problem		
3658	2015	All other reportable incidents			CT257	Implantable cardiac pacemakers			A13	Communication or Transmission Problem		
									A26	Insufficient Device Problem Information		
3659	2016	Unanticipated serious deterioration in state of health			CT1684	Cardiac pacemakers and associated devices			A15	Activation, Positioning or Separation Problem		
									A150203	Difficult or Delayed Positioning		
3660	2015	All other reportable incidents			CT257	Implantable cardiac pacemakers			A07	Electrical /Electronic Property Problem	06/02/2015	
									A0708	Power Problem	06/02/2015	
3661	2015				CT257	Implantable cardiac pacemakers			A26	Insufficient Device Problem Information		

3662	2015		47270	Cardiac resynchronization therapy implantable defibrillator	CT2866	Implantable cardiac electrical stimulators			A26	Insufficient Device Problem Information		
3663	2015		47270	Cardiac resynchronization therapy implantable defibrillator	CT2866	Implantable cardiac electrical stimulators			A26	Insufficient Device Problem Information		
3664	2015				CT257	Implantable cardiac pacemakers			A26	Insufficient Device Problem Information		
3665	2015		47270	Cardiac resynchronization therapy implantable defibrillator	CT2866	Implantable cardiac electrical stimulators			A07	Electrical /Electronic Property Problem		
									A0708	Power Problem		
3666	2015		47270	Cardiac resynchronization therapy implantable defibrillator	CT2866	Implantable cardiac electrical stimulators			A07	Electrical /Electronic Property Problem		
									A0708	Power Problem		
3667	2016	All other reportable incidents	47270	Cardiac resynchronization therapy implantable defibrillator	CT2866	Implantable cardiac electrical stimulators			A07	Electrical /Electronic Property Problem	21/09/2016	
									A0708	Power Problem	21/09/2016	
3668	2016	All other reportable incidents	47263	Cardiac resynchronization therapy implantable pacemaker	CT1684	Cardiac pacemakers and associated devices			A07	Electrical /Electronic Property Problem	10/12/2015	
									A0708	Power Problem	10/12/2015	
					CT1685	Cardiac pacemakers			A07	Electrical /Electronic Property Problem	10/12/2015	
									A0708	Power Problem	10/12/2015	
					CT257	Implantable cardiac pacemakers			A07	Electrical /Electronic Property Problem	10/12/2015	
									A0708	Power Problem	10/12/2015	
					CT2866	Implantable cardiac electrical stimulators			A07	Electrical /Electronic Property Problem	10/12/2015	
									A0708	Power Problem	10/12/2015	
3669	2015	All other reportable incidents	47270	Cardiac resynchronization therapy implantable defibrillator	CT2866	Implantable cardiac electrical stimulators			A13	Communication or Transmission Problem		
									A1304	Telemetry Discrepancy		
									A26	Insufficient Device Problem Information		
3670	2016	TBD	35223	Endocardial/interventricular septal pacing lead	CT1684	Cardiac pacemakers and associated devices			A26	Insufficient Device Problem Information		
3671	2016	Death			CT257	Implantable cardiac pacemakers			A26	Insufficient Device Problem Information	28/04/2016	
3672	2015		35223	Endocardial/interventricular septal pacing lead	CT1684	Cardiac pacemakers and associated devices			A0104	Migration or Expulsion of Device		
3673	2015		47270	Cardiac resynchronization therapy implantable defibrillator	CT2866	Implantable cardiac electrical stimulators			A26	Insufficient Device Problem Information		
3674	2015		47270	Cardiac resynchronization therapy implantable defibrillator	CT2866	Implantable cardiac electrical stimulators			A26	Insufficient Device Problem Information		
3675	2016	All other reportable incidents	47263	Cardiac resynchronization therapy implantable pacemaker	CT1684	Cardiac pacemakers and associated devices			A07	Electrical /Electronic Property Problem	13/09/2016	
									A0709	Device Sensing Problem	13/09/2016	
					CT1685	Cardiac pacemakers			A07	Electrical /Electronic Property Problem	13/09/2016	
									A0709	Device Sensing Problem	13/09/2016	
					CT257	Implantable cardiac pacemakers			A07	Electrical /Electronic Property Problem	13/09/2016	
									A0709	Device Sensing Problem	13/09/2016	
					CT2866	Implantable cardiac electrical stimulators			A07	Electrical /Electronic Property Problem	13/09/2016	
									A0709	Device Sensing Problem	13/09/2016	
3676	2016	TBD	47270	Cardiac resynchronization therapy implantable defibrillator	CT2866	Implantable cardiac electrical stimulators			A07	Electrical /Electronic Property Problem		
									A0708	Power Problem		
3677	2016	TBD	35223	Endocardial/interventricular septal pacing lead	CT1684	Cardiac pacemakers and associated devices			A26	Insufficient Device Problem Information		
3678	2015		47270	Cardiac resynchronization therapy implantable defibrillator	CT2866	Implantable cardiac electrical stimulators			A26	Insufficient Device Problem Information		
3679	2015		47270	Cardiac resynchronization therapy implantable defibrillator	CT2866	Implantable cardiac electrical stimulators			A26	Insufficient Device Problem Information		
3680	2016	TBD			CT257	Implantable cardiac pacemakers			A07	Electrical /Electronic Property Problem		
									A0709	Device Sensing Problem		
3681	2015				CT257	Implantable cardiac pacemakers			A26	Insufficient Device Problem Information		
3682	2016	TBD	47263	Cardiac resynchronization therapy implantable pacemaker	CT1684	Cardiac pacemakers and associated devices			A07	Electrical /Electronic Property Problem	25/05/2016	
									A0708	Power Problem	25/05/2016	
									A0709	Device Sensing Problem	25/05/2016	
					CT1685	Cardiac pacemakers			A07	Electrical /Electronic Property Problem	25/05/2016	
									A0708	Power Problem	25/05/2016	
									A0709	Device Sensing Problem	25/05/2016	
					CT257	Implantable cardiac pacemakers			A07	Electrical /Electronic Property Problem	25/05/2016	
									A0708	Power Problem	25/05/2016	
									A0709	Device Sensing Problem	25/05/2016	
					CT2866	Implantable cardiac electrical stimulators			A07	Electrical /Electronic Property Problem	25/05/2016	
									A0708	Power Problem	25/05/2016	
									A0709	Device Sensing Problem	25/05/2016	
3683	2015				CT257	Implantable cardiac pacemakers			A07	Electrical /Electronic Property Problem		
									A0708	Power Problem		
									A13	Communication or Transmission Problem		
									A1304	Telemetry Discrepancy		
									A26	Insufficient Device Problem Information		
3684	2016	All other reportable incidents			CT257	Implantable cardiac pacemakers			A07	Electrical /Electronic Property Problem		
									A0708	Power Problem		
3685	2016	All other reportable incidents			CT257	Implantable cardiac pacemakers			A13	Communication or Transmission Problem	18/04/2016	
									A26	Insufficient Device Problem Information	18/04/2016	
3686	2016	TBD			CT257	Implantable cardiac pacemakers			A07	Electrical /Electronic Property Problem		
									A0708	Power Problem		
3687	2015				CT257	Implantable cardiac pacemakers			A26	Insufficient Device Problem Information		
3688	2016	TBD	35223	Endocardial/interventricular septal pacing lead	CT1684	Cardiac pacemakers and associated devices			A07	Electrical /Electronic Property Problem		
									A0709	Device Sensing Problem		
3689	2015				CT257	Implantable cardiac pacemakers			A26	Insufficient Device Problem Information		
3690	2015		47270	Cardiac resynchronization therapy implantable defibrillator	CT2866	Implantable cardiac electrical stimulators			A07	Electrical /Electronic Property Problem		
									A0708	Power Problem		
3691	2016	All other reportable incidents	47270	Cardiac resynchronization therapy implantable defibrillator	CT2866	Implantable cardiac electrical stimulators			A07	Electrical /Electronic Property Problem		
									A0721	Circuit Failure		
3692	2015				CT257	Implantable cardiac pacemakers			A26	Insufficient Device Problem Information		
3693	2015		47270	Cardiac resynchronization therapy implantable defibrillator	CT2866	Implantable cardiac electrical stimulators			A26	Insufficient Device Problem Information		
3694	2016	TBD	35852	Single-chamber implantable defibrillator	CT2866	Implantable cardiac electrical stimulators			A07	Electrical /Electronic Property Problem		
									A0708	Power Problem		
3695	2015	All other reportable incidents	47270	Cardiac resynchronization therapy implantable defibrillator	CT2866	Implantable cardiac electrical stimulators			A07	Electrical /Electronic Property Problem	14/07/2015	
									A0708	Power Problem	14/07/2015	
3696	2015	All other reportable incidents	47270	Cardiac resynchronization therapy implantable defibrillator	CT2866	Implantable cardiac electrical stimulators			A09	Output Problem		
									A090404	Output Above Specifications		

3697	2015	All other reportable incidents			CT257	Implantable cardiac pacemakers			A11	Computer Software Problem	13/02/2015	
									A1102	Application Program Problem	13/02/2015	
3698	2016	TBD	47270	Cardiac resynchronization therapy implantable defibrillator	CT2866	Implantable cardiac electrical stimulators			A13	Communication or Transmission Problem	03/05/2016	
									A26	Insufficient Device Problem Information	03/05/2016	
3699	2015	All other reportable incidents			CT257	Implantable cardiac pacemakers			A16	Protective Measures Problem	01/10/2015	
									A1602	Fail-Safe Problem	01/10/2015	
3700	2016	TBD	47270	Cardiac resynchronization therapy implantable defibrillator	CT2866	Implantable cardiac electrical stimulators			A07	Electrical /Electronic Property Problem		
									A0708	Power Problem		
3701	2016	All other reportable incidents	47270	Cardiac resynchronization therapy implantable defibrillator	CT2866	Implantable cardiac electrical stimulators			A07	Electrical /Electronic Property Problem	29/04/2016	
									A0703	Arcing	29/04/2016	
3702	2016	All other reportable incidents	47263	Cardiac resynchronization therapy implantable pacemaker	CT1684	Cardiac pacemakers and associated devices			A07	Electrical /Electronic Property Problem	14/12/2015	
									A0708	Power Problem	14/12/2015	
					CT1685	Cardiac pacemakers			A07	Electrical /Electronic Property Problem	14/12/2015	
									A0708	Power Problem	14/12/2015	
					CT257	Implantable cardiac pacemakers			A07	Electrical /Electronic Property Problem	14/12/2015	
									A0708	Power Problem	14/12/2015	
					CT2866	Implantable cardiac electrical stimulators			A07	Electrical /Electronic Property Problem	14/12/2015	
									A0708	Power Problem	14/12/2015	
3703	2015		47270	Cardiac resynchronization therapy implantable defibrillator	CT2866	Implantable cardiac electrical stimulators			A26	Insufficient Device Problem Information		
3704	2016	All other reportable incidents			CT257	Implantable cardiac pacemakers			A07	Electrical /Electronic Property Problem	01/10/2016	
									A0708	Power Problem	01/10/2016	
3705	2015		47270	Cardiac resynchronization therapy implantable defibrillator	CT2866	Implantable cardiac electrical stimulators			A26	Insufficient Device Problem Information		
3706	2015				CT257	Implantable cardiac pacemakers			A07	Electrical /Electronic Property Problem		
									A0708	Power Problem		
3707	2016	TBD	35223	Endocardial/interventricular septal pacing lead	CT1684	Cardiac pacemakers and associated devices			A26	Insufficient Device Problem Information		
3708	2016	All other reportable incidents			CT1684	Cardiac pacemakers and associated devices			A12	Connection Problem		
									A1203	Disconnection		
3709	2015		47270	Cardiac resynchronization therapy implantable defibrillator	CT2866	Implantable cardiac electrical stimulators			A26	Insufficient Device Problem Information		
3710	2016	All other reportable incidents			CT257	Implantable cardiac pacemakers			A07	Electrical /Electronic Property Problem	11/10/2016	
									A0708	Power Problem	11/10/2016	
3711	2015	All other reportable incidents			CT257	Implantable cardiac pacemakers			A07	Electrical /Electronic Property Problem		
									A0708	Power Problem		
3712	2015	All other reportable incidents	47270	Cardiac resynchronization therapy implantable defibrillator	CT2866	Implantable cardiac electrical stimulators			A07	Electrical /Electronic Property Problem		
									A0708	Power Problem		
3713	2015		47270	Cardiac resynchronization therapy implantable defibrillator	CT2866	Implantable cardiac electrical stimulators			A26	Insufficient Device Problem Information		
3714	2016	TBD			CT1684	Cardiac pacemakers and associated devices			A09	Output Problem		
									A0907	No Device Output		
3715	2015		47270	Cardiac resynchronization therapy implantable defibrillator	CT2866	Implantable cardiac electrical stimulators			A26	Insufficient Device Problem Information		
3716	2015		35223	Endocardial/interventricular septal pacing lead	CT1684	Cardiac pacemakers and associated devices			A07	Electrical /Electronic Property Problem		
									A0709	Device Sensing Problem		
3717	2015		47270	Cardiac resynchronization therapy implantable defibrillator	CT2866	Implantable cardiac electrical stimulators			A07	Electrical /Electronic Property Problem		
									A0708	Power Problem		
3718	2015		47270	Cardiac resynchronization therapy implantable defibrillator	CT2866	Implantable cardiac electrical stimulators			A07	Electrical /Electronic Property Problem		
									A0708	Power Problem		
3719	2015	All other reportable incidents			CT1684	Cardiac pacemakers and associated devices			A12	Connection Problem		
3720	2016	TBD			CT257	Implantable cardiac pacemakers			A07	Electrical /Electronic Property Problem	13/10/2016	
									A0708	Power Problem	13/10/2016	
3721	2016	All other reportable incidents	47270	Cardiac resynchronization therapy implantable defibrillator	CT2866	Implantable cardiac electrical stimulators			A07	Electrical /Electronic Property Problem	06/05/2016	
									A0708	Power Problem	06/05/2016	
3722	2015				CT257	Implantable cardiac pacemakers			A26	Insufficient Device Problem Information		
3723	2015				CT257	Implantable cardiac pacemakers			A1304	Telemetry Discrepancy		
									A26	Insufficient Device Problem Information		
3724	2015				CT257	Implantable cardiac pacemakers			A11	Computer Software Problem		
									A1102	Application Program Problem		
3725	2015		35852	Single-chamber implantable defibrillator	CT2866	Implantable cardiac electrical stimulators			A12	Connection Problem		
									A1208	Fitting Problem		
3726	2015		47270	Cardiac resynchronization therapy implantable defibrillator	CT2866	Implantable cardiac electrical stimulators			A07	Electrical /Electronic Property Problem		
									A0708	Power Problem		
3727	2015	All other reportable incidents	47270	Cardiac resynchronization therapy implantable defibrillator	CT2866	Implantable cardiac electrical stimulators			A16	Protective Measures Problem		
									A1602	Fail-Safe Problem		
3728	2015		47270	Cardiac resynchronization therapy implantable defibrillator	CT2866	Implantable cardiac electrical stimulators			A26	Insufficient Device Problem Information		
3729	2015		47270	Cardiac resynchronization therapy implantable defibrillator	CT2866	Implantable cardiac electrical stimulators			A07	Electrical /Electronic Property Problem		
									A0708	Power Problem		
3730	2016	All other reportable incidents			CT257	Implantable cardiac pacemakers			A07	Electrical /Electronic Property Problem		
									A0709	Device Sensing Problem		
3731	2016	All other reportable incidents			CT257	Implantable cardiac pacemakers			A07	Electrical /Electronic Property Problem	16/09/2016	
									A0708	Power Problem	16/09/2016	
									A13	Communication or Transmission Problem	16/09/2016	
									A26	Insufficient Device Problem Information	16/09/2016	
3732	2015		47270	Cardiac resynchronization therapy implantable defibrillator	CT2866	Implantable cardiac electrical stimulators			A07	Electrical /Electronic Property Problem		
									A0708	Power Problem		
3733	2016	TBD	35223	Endocardial/interventricular septal pacing lead	CT1684	Cardiac pacemakers and associated devices			A26	Insufficient Device Problem Information		
3734	2016	All other reportable incidents			CT257	Implantable cardiac pacemakers			A07	Electrical /Electronic Property Problem	13/10/2016	
									A0708	Power Problem	13/10/2016	
									A09	Output Problem	13/10/2016	
									A0907	No Device Output	13/10/2016	
3735	2015				CT257	Implantable cardiac pacemakers			A26	Insufficient Device Problem Information		
3736	2015	All other reportable incidents	47270	Cardiac resynchronization therapy implantable defibrillator	CT2866	Implantable cardiac electrical stimulators			A07	Electrical /Electronic Property Problem		
									A0708	Power Problem		
3737	2015		47270	Cardiac resynchronization therapy implantable defibrillator	CT2866	Implantable cardiac electrical stimulators			A26	Insufficient Device Problem Information		
3738	2015				CT257	Implantable cardiac pacemakers			A26	Insufficient Device Problem Information		

3739	2016	TBD	35223	Endocardial/interventricular septal pacing lead	CT1684	Cardiac pacemakers and associated devices			A07	Electrical /Electronic Property Problem	16/05/2016	
									A0709	Device Sensing Problem	16/05/2016	
3740	2016	TBD	47270	Cardiac resynchronization therapy implantable defibrillator	CT2866	Implantable cardiac electrical stimulators			A07	Electrical /Electronic Property Problem		
									A0708	Power Problem		
3741	2016	TBD	47270	Cardiac resynchronization therapy implantable defibrillator	CT2866	Implantable cardiac electrical stimulators			A07	Electrical /Electronic Property Problem		
									A0708	Power Problem		
3742	2016	TBD			CT257	Implantable cardiac pacemakers			A07	Electrical /Electronic Property Problem		
									A0709	Device Sensing Problem		
3743	2015				CT257	Implantable cardiac pacemakers			A26	Insufficient Device Problem Information		
3744	2015	All other reportable incidents	47270	Cardiac resynchronization therapy implantable defibrillator	CT2866	Implantable cardiac electrical stimulators			A07	Electrical /Electronic Property Problem		
									A0709	Device Sensing Problem		
									A09	Output Problem		
									A090404	Output Above Specifications		
3745	2015		47270	Cardiac resynchronization therapy implantable defibrillator	CT2866	Implantable cardiac electrical stimulators			A26	Insufficient Device Problem Information		
3746	2016	TBD	35223	Endocardial/interventricular septal pacing lead	CT1684	Cardiac pacemakers and associated devices			A26	Insufficient Device Problem Information		
3747	2015	All other reportable incidents			CT257	Implantable cardiac pacemakers			A07	Electrical /Electronic Property Problem	04/12/2015	
									A0708	Power Problem	04/12/2015	
3748	2016	TBD			CT257	Implantable cardiac pacemakers			A07	Electrical /Electronic Property Problem		
									A0708	Power Problem		
3749	2015				CT1684	Cardiac pacemakers and associated devices			A11	Computer Software Problem		
									A1102	Application Program Problem		
3750	2015				CT257	Implantable cardiac pacemakers			A07	Electrical /Electronic Property Problem		
									A0708	Power Problem		
3751	2015		47270	Cardiac resynchronization therapy implantable defibrillator	CT2866	Implantable cardiac electrical stimulators			A26	Insufficient Device Problem Information		
3752	2015				CT257	Implantable cardiac pacemakers			A26	Insufficient Device Problem Information		
3753	2015		47270	Cardiac resynchronization therapy implantable defibrillator	CT2866	Implantable cardiac electrical stimulators			A07	Electrical /Electronic Property Problem		
									A0708	Power Problem		
3754	2015				CT257	Implantable cardiac pacemakers			A07	Electrical /Electronic Property Problem		
									A0708	Power Problem		
3755	2016	TBD	47263	Cardiac resynchronization therapy implantable pacemaker	CT1684	Cardiac pacemakers and associated devices			A07	Electrical /Electronic Property Problem	19/05/2016	
									A0708	Power Problem	19/05/2016	
					CT1685	Cardiac pacemakers			A07	Electrical /Electronic Property Problem	19/05/2016	
									A0708	Power Problem	19/05/2016	
					CT257	Implantable cardiac pacemakers			A07	Electrical /Electronic Property Problem	19/05/2016	
									A0708	Power Problem	19/05/2016	
					CT2866	Implantable cardiac electrical stimulators			A07	Electrical /Electronic Property Problem	19/05/2016	
									A0708	Power Problem	19/05/2016	
3756	2015		47270	Cardiac resynchronization therapy implantable defibrillator	CT2866	Implantable cardiac electrical stimulators			A07	Electrical /Electronic Property Problem		
									A0708	Power Problem		
3757	2015				CT257	Implantable cardiac pacemakers			A1304	Telemetry Discrepancy		
									A26	Insufficient Device Problem Information		
3758	2016	All other reportable incidents	47270	Cardiac resynchronization therapy implantable defibrillator	CT2866	Implantable cardiac electrical stimulators			A09	Output Problem		
									A0907	No Device Output		
3759	2015	All other reportable incidents	47270	Cardiac resynchronization therapy implantable defibrillator	CT2866	Implantable cardiac electrical stimulators			A07	Electrical /Electronic Property Problem	31/01/2016	
									A0721	Circuit Failure	31/01/2016	
3760	2016	All other reportable incidents	47270	Cardiac resynchronization therapy implantable defibrillator	CT2866	Implantable cardiac electrical stimulators			A07	Electrical /Electronic Property Problem		
									A0708	Power Problem		
3761	2016	TBD	47270	Cardiac resynchronization therapy implantable defibrillator	CT2866	Implantable cardiac electrical stimulators			A07	Electrical /Electronic Property Problem		
									A0708	Power Problem		
3762	2015		47270	Cardiac resynchronization therapy implantable defibrillator	CT2866	Implantable cardiac electrical stimulators			A07	Electrical /Electronic Property Problem		
									A0708	Power Problem		
3763	2015	All other reportable incidents			CT257	Implantable cardiac pacemakers			A07	Electrical /Electronic Property Problem		
									A0709	Device Sensing Problem		
3764	2015				CT257	Implantable cardiac pacemakers			A26	Insufficient Device Problem Information		
3765	2016	All other reportable incidents			CT257	Implantable cardiac pacemakers			A07	Electrical /Electronic Property Problem		
									A0709	Device Sensing Problem		
3766	2016	TBD	35223	Endocardial/interventricular septal pacing lead	CT1684	Cardiac pacemakers and associated devices			A26	Insufficient Device Problem Information		
3767	2016	All other reportable incidents	47270	Cardiac resynchronization therapy implantable defibrillator	CT2866	Implantable cardiac electrical stimulators			A07	Electrical /Electronic Property Problem	16/05/2016	
									A0703	Arcing	16/05/2016	
3768	2015				CT257	Implantable cardiac pacemakers			A13	Communication or Transmission Problem		
									A26	Insufficient Device Problem Information		
3769	2015				CT257	Implantable cardiac pacemakers			A07	Electrical /Electronic Property Problem	08/07/2015	
									A0708	Power Problem	08/07/2015	
3770	2016	TBD	47270	Cardiac resynchronization therapy implantable defibrillator	CT2866	Implantable cardiac electrical stimulators			A07	Electrical /Electronic Property Problem	21/10/2016	
									A0708	Power Problem	21/10/2016	
3771	2016	All other reportable incidents	47270	Cardiac resynchronization therapy implantable defibrillator	CT2866	Implantable cardiac electrical stimulators			A07	Electrical /Electronic Property Problem	31/10/2016	
									A0708	Power Problem	31/10/2016	
3772	2015				CT257	Implantable cardiac pacemakers			A26	Insufficient Device Problem Information		
3773	2015		47270	Cardiac resynchronization therapy implantable defibrillator	CT2866	Implantable cardiac electrical stimulators			A07	Electrical /Electronic Property Problem		
									A0708	Power Problem		
3774	2016	All other reportable incidents	47263	Cardiac resynchronization therapy implantable pacemaker	CT1684	Cardiac pacemakers and associated devices			A07	Electrical /Electronic Property Problem	10/08/2016	
									A0708	Power Problem	10/08/2016	
					CT1685	Cardiac pacemakers			A07	Electrical /Electronic Property Problem	10/08/2016	
									A0708	Power Problem	10/08/2016	
					CT257	Implantable cardiac pacemakers			A07	Electrical /Electronic Property Problem	10/08/2016	
									A0708	Power Problem	10/08/2016	
					CT2866	Implantable cardiac electrical stimulators			A07	Electrical /Electronic Property Problem	10/08/2016	
									A0708	Power Problem	10/08/2016	
3775	2015	All other reportable incidents			CT257	Implantable cardiac pacemakers			A07	Electrical /Electronic Property Problem		
									A0708	Power Problem		
3776	2015	All other reportable incidents			CT257	Implantable cardiac pacemakers			A07	Electrical /Electronic Property Problem		
									A0708	Power Problem		
3777	2016	TBD	35223	Endocardial/interventricular septal pacing lead	CT1684	Cardiac pacemakers and associated devices			A26	Insufficient Device Problem Information		
3778	2016	Death	35223	Endocardial/interventricular septal pacing lead	CT1684	Cardiac pacemakers and associated devices			A07	Electrical /Electronic Property Problem		

									A0709	Device Sensing Problem		
3779	2015	All other reportable incidents			CT257	Implantable cardiac pacemakers			A07	Electrical /Electronic Property Problem		
									A0708	Power Problem		
3780	2015		47270	Cardiac resynchronization therapy implantable defibrillator	CT2866	Implantable cardiac electrical stimulators			A26	Insufficient Device Problem Information		
3781	2016	All other reportable incidents	47263	Cardiac resynchronization therapy implantable pacemaker	CT1684	Cardiac pacemakers and associated devices			A16	Protective Measures Problem		
									A1602	Fail-Safe Problem		
					CT1685	Cardiac pacemakers			A16	Protective Measures Problem		
									A1602	Fail-Safe Problem		
					CT257	Implantable cardiac pacemakers			A16	Protective Measures Problem		
									A1602	Fail-Safe Problem		
					CT2866	Implantable cardiac electrical stimulators			A16	Protective Measures Problem		
									A1602	Fail-Safe Problem		
3782	2015				CT257	Implantable cardiac pacemakers			A26	Insufficient Device Problem Information		
3783	2016	TBD			CT257	Implantable cardiac pacemakers			A07	Electrical /Electronic Property Problem	24/10/2016	
									A0708	Power Problem	24/10/2016	
3784	2015		47270	Cardiac resynchronization therapy implantable defibrillator	CT2866	Implantable cardiac electrical stimulators			A07	Electrical /Electronic Property Problem		
									A0708	Power Problem		
3785	2015				CT257	Implantable cardiac pacemakers			A07	Electrical /Electronic Property Problem		
									A0709	Device Sensing Problem		
3786	2015		47270	Cardiac resynchronization therapy implantable defibrillator	CT2866	Implantable cardiac electrical stimulators			A07	Electrical /Electronic Property Problem		
									A0708	Power Problem		
3787	2016	All other reportable incidents	47270	Cardiac resynchronization therapy implantable defibrillator	CT2866	Implantable cardiac electrical stimulators			A07	Electrical /Electronic Property Problem		
									A0708	Power Problem		
3788	2016	TBD			CT257	Implantable cardiac pacemakers			A26	Insufficient Device Problem Information		
3789	2016	TBD	35223	Endocardial/interventricular septal pacing lead	CT1684	Cardiac pacemakers and associated devices			A26	Insufficient Device Problem Information		
3790	2015	All other reportable incidents			CT257	Implantable cardiac pacemakers			A07	Electrical /Electronic Property Problem		
									A0708	Power Problem		
3791	2016	Unanticipated serious deterioration in state of health	35852	Single-chamber implantable defibrillator	CT2866	Implantable cardiac electrical stimulators			A01	Patient Device Interaction Problem	20/05/2016	
									A17	Compatibility Problem	20/05/2016	
3792	2015		47270	Cardiac resynchronization therapy implantable defibrillator	CT2866	Implantable cardiac electrical stimulators			A26	Insufficient Device Problem Information		
3793	2015		47270	Cardiac resynchronization therapy implantable defibrillator	CT2866	Implantable cardiac electrical stimulators			A26	Insufficient Device Problem Information		
3794	2015				CT257	Implantable cardiac pacemakers			A26	Insufficient Device Problem Information		
3795	2015		47270	Cardiac resynchronization therapy implantable defibrillator	CT2866	Implantable cardiac electrical stimulators			A26	Insufficient Device Problem Information		
3796	2016	TBD			CT257	Implantable cardiac pacemakers			A07	Electrical /Electronic Property Problem		
									A0708	Power Problem		
3797	2016	TBD			CT257	Implantable cardiac pacemakers			A07	Electrical /Electronic Property Problem		
									A0708	Power Problem		
3798	2015		47270	Cardiac resynchronization therapy implantable defibrillator	CT2866	Implantable cardiac electrical stimulators			A07	Electrical /Electronic Property Problem		
									A0709	Device Sensing Problem		
									A09	Output Problem		
									A090404	Output Above Specifications		
3799	2015				CT257	Implantable cardiac pacemakers			A1304	Telemetry Discrepancy		
									A26	Insufficient Device Problem Information		
3800	2015				CT1684	Cardiac pacemakers and associated devices			A09	Output Problem		
									A0907	No Device Output		
3801	2016	TBD	47270	Cardiac resynchronization therapy implantable defibrillator	CT2866	Implantable cardiac electrical stimulators			A07	Electrical /Electronic Property Problem		
									A0709	Device Sensing Problem		
3802	2016	TBD			CT257	Implantable cardiac pacemakers			A15	Activation, Positioning or Separation Problem		
									A150203	Difficult or Delayed Positioning		
3803	2015				CT257	Implantable cardiac pacemakers			A07	Electrical /Electronic Property Problem	15/07/2015	
									A0708	Power Problem	15/07/2015	
3804	2016	TBD	47270	Cardiac resynchronization therapy implantable defibrillator	CT2866	Implantable cardiac electrical stimulators			A07	Electrical /Electronic Property Problem		
									A0709	Device Sensing Problem		
3805	2015				CT257	Implantable cardiac pacemakers			A07	Electrical /Electronic Property Problem	23/07/2015	
									A0708	Power Problem	23/07/2015	
3806	2016	All other reportable incidents	47263	Cardiac resynchronization therapy implantable pacemaker	CT1684	Cardiac pacemakers and associated devices			A07	Electrical /Electronic Property Problem	28/09/2016	
									A0708	Power Problem	28/09/2016	
					CT1685	Cardiac pacemakers			A07	Electrical /Electronic Property Problem	28/09/2016	
									A0708	Power Problem	28/09/2016	
					CT257	Implantable cardiac pacemakers			A07	Electrical /Electronic Property Problem	28/09/2016	
									A0708	Power Problem	28/09/2016	
					CT2866	Implantable cardiac electrical stimulators			A07	Electrical /Electronic Property Problem	28/09/2016	
									A0708	Power Problem	28/09/2016	
3807	2016	TBD	47270	Cardiac resynchronization therapy implantable defibrillator	CT2866	Implantable cardiac electrical stimulators			A07	Electrical /Electronic Property Problem		
									A0708	Power Problem		
3808	2016	All other reportable incidents	47270	Cardiac resynchronization therapy implantable defibrillator	CT2866	Implantable cardiac electrical stimulators			A07	Electrical /Electronic Property Problem		
									A0708	Power Problem		
3809	2015		47270	Cardiac resynchronization therapy implantable defibrillator	CT2866	Implantable cardiac electrical stimulators			A07	Electrical /Electronic Property Problem		
									A0708	Power Problem		
3810	2016	All other reportable incidents	47270	Cardiac resynchronization therapy implantable defibrillator	CT2866	Implantable cardiac electrical stimulators			A13	Communication or Transmission Problem	03/05/2016	
									A26	Insufficient Device Problem Information	03/05/2016	
3811	2016	TBD	35223	Endocardial/interventricular septal pacing lead	CT1684	Cardiac pacemakers and associated devices			A26	Insufficient Device Problem Information		
3812	2016	All other reportable incidents	47270	Cardiac resynchronization therapy implantable defibrillator	CT2866	Implantable cardiac electrical stimulators			A07	Electrical /Electronic Property Problem	15/12/2015	
									A0708	Power Problem	15/12/2015	
									A1104	Computer Operating System Problem	15/12/2015	
3813	2016	All other reportable incidents	35223	Endocardial/interventricular septal pacing lead	CT1684	Cardiac pacemakers and associated devices			A0104	Migration or Expulsion of Device	17/10/2016	
3814	2015	All other reportable incidents			CT257	Implantable cardiac pacemakers			A07	Electrical /Electronic Property Problem		
									A0708	Power Problem		
3815	2015		47270	Cardiac resynchronization therapy implantable defibrillator	CT2866	Implantable cardiac electrical stimulators			A26	Insufficient Device Problem Information		
3816	2016	TBD			CT1684	Cardiac pacemakers and associated devices			A09	Output Problem		
									A0908	Incorrect, Inadequate or Imprecise Result or Readings		

3817	2016	TBD			CT257	Implantable cardiac pacemakers			A13	Communication or Transmission Problem		
									A26	Insufficient Device Problem Information		
3818	2015	All other reportable incidents			CT1684	Cardiac pacemakers and associated devices			A17	Compatibility Problem		
									A1701	Component or Accessory Incompatibility		
3819	2016	TBD			CT257	Implantable cardiac pacemakers			A1104	Computer Operating System Problem		
3820	2016	TBD	47270	Cardiac resynchronization therapy implantable defibrillator	CT2866	Implantable cardiac electrical stimulators			A11	Computer Software Problem		
									A1102	Application Program Problem		
3821	2016	TBD			CT257	Implantable cardiac pacemakers			A07	Electrical /Electronic Property Problem		
									A0708	Power Problem		
3822	2016	TBD			CT257	Implantable cardiac pacemakers			A13	Communication or Transmission Problem		
									A26	Insufficient Device Problem Information		
3823	2015		47270	Cardiac resynchronization therapy implantable defibrillator	CT2866	Implantable cardiac electrical stimulators			A07	Electrical /Electronic Property Problem		
									A0708	Power Problem		
3824	2016	All other reportable incidents	47270	Cardiac resynchronization therapy implantable defibrillator	CT2866	Implantable cardiac electrical stimulators			A1104	Computer Operating System Problem		
3825	2016	TBD	47270	Cardiac resynchronization therapy implantable defibrillator	CT2866	Implantable cardiac electrical stimulators			A13	Communication or Transmission Problem		
									A26	Insufficient Device Problem Information		
3826	2015		47270	Cardiac resynchronization therapy implantable defibrillator	CT2866	Implantable cardiac electrical stimulators			A26	Insufficient Device Problem Information		
3827	2015		35223	Endocardial/interventricular septal pacing lead	CT1684	Cardiac pacemakers and associated devices			A04	Material Integrity Problem		
									A0404	Crack		
3828	2016	TBD	47270	Cardiac resynchronization therapy implantable defibrillator	CT2866	Implantable cardiac electrical stimulators			A07	Electrical /Electronic Property Problem		
									A0708	Power Problem		
3829	2015	All other reportable incidents	47263	Cardiac resynchronization therapy implantable pacemaker	CT1684	Cardiac pacemakers and associated devices			A07	Electrical /Electronic Property Problem		
									A0708	Power Problem		
					CT1685	Cardiac pacemakers			A07	Electrical /Electronic Property Problem		
									A0708	Power Problem		
					CT257	Implantable cardiac pacemakers			A07	Electrical /Electronic Property Problem		
									A0708	Power Problem		
					CT2866	Implantable cardiac electrical stimulators			A07	Electrical /Electronic Property Problem		
									A0708	Power Problem		
3830	2016	TBD			CT257	Implantable cardiac pacemakers			A26	Insufficient Device Problem Information		
3831	2016	TBD	35223	Endocardial/interventricular septal pacing lead	CT1684	Cardiac pacemakers and associated devices			A05	Mechanical Problem		
									A0512	Unintended Movement		
3832	2015		47270	Cardiac resynchronization therapy implantable defibrillator	CT2866	Implantable cardiac electrical stimulators			A26	Insufficient Device Problem Information		
3833	2016	TBD	35223	Endocardial/interventricular septal pacing lead	CT1684	Cardiac pacemakers and associated devices			A15	Activation, Positioning or Separation Problem		
									A150203	Difficult or Delayed Positioning		
3834	2015				CT1684	Cardiac pacemakers and associated devices			A15	Activation, Positioning or Separation Problem		
									A150101	Activation Failure		
3835	2015				CT257	Implantable cardiac pacemakers			A11	Computer Software Problem		
									A1102	Application Program Problem		
3836	2015		47270	Cardiac resynchronization therapy implantable defibrillator	CT2866	Implantable cardiac electrical stimulators			A26	Insufficient Device Problem Information		
3837	2016	All other reportable incidents	47270	Cardiac resynchronization therapy implantable defibrillator	CT2866	Implantable cardiac electrical stimulators			A07	Electrical /Electronic Property Problem	09/12/2015	
									A0708	Power Problem	09/12/2015	
3838	2016	All other reportable incidents	47270	Cardiac resynchronization therapy implantable defibrillator	CT2866	Implantable cardiac electrical stimulators			A07	Electrical /Electronic Property Problem	04/10/2016	
									A0708	Power Problem	04/10/2016	
3839	2016	All other reportable incidents			CT257	Implantable cardiac pacemakers			A07	Electrical /Electronic Property Problem	05/05/2016	
									A0708	Power Problem	05/05/2016	
3840	2016	TBD			CT257	Implantable cardiac pacemakers			A07	Electrical /Electronic Property Problem		
									A0709	Device Sensing Problem		
									A09	Output Problem		
									A0907	No Device Output		
3841	2015	All other reportable incidents			CT257	Implantable cardiac pacemakers			A09	Output Problem		
									A0907	No Device Output		
									A1304	Telemetry Discrepancy		
									A26	Insufficient Device Problem Information		
3842	2015		47270	Cardiac resynchronization therapy implantable defibrillator	CT2866	Implantable cardiac electrical stimulators			A26	Insufficient Device Problem Information		
3843	2016	TBD			CT1684	Cardiac pacemakers and associated devices			A11	Computer Software Problem		
									A1102	Application Program Problem		
3844	2015				CT257	Implantable cardiac pacemakers			A26	Insufficient Device Problem Information		
3845	2016	TBD			CT257	Implantable cardiac pacemakers			A26	Insufficient Device Problem Information		
3846	2016	TBD	47270	Cardiac resynchronization therapy implantable defibrillator	CT2866	Implantable cardiac electrical stimulators			A07	Electrical /Electronic Property Problem		
									A0708	Power Problem		
3847	2016	TBD	35223	Endocardial/interventricular septal pacing lead	CT1684	Cardiac pacemakers and associated devices			A26	Insufficient Device Problem Information		
3848	2016	All other reportable incidents			CT1684	Cardiac pacemakers and associated devices			A09	Output Problem		
									A0907	No Device Output		
3849	2016	TBD			CT257	Implantable cardiac pacemakers			A07	Electrical /Electronic Property Problem		
									A0708	Power Problem		
3850	2015				CT257	Implantable cardiac pacemakers			A07	Electrical /Electronic Property Problem		
									A0708	Power Problem		
3851	2016	All other reportable incidents			CT257	Implantable cardiac pacemakers			A07	Electrical /Electronic Property Problem		
									A0708	Power Problem		
									A0709	Device Sensing Problem		
3852	2016	All other reportable incidents	47270	Cardiac resynchronization therapy implantable defibrillator	CT2866	Implantable cardiac electrical stimulators			A13	Communication or Transmission Problem	18/12/2015	
									A26	Insufficient Device Problem Information	18/12/2015	
3853	2015				CT257	Implantable cardiac pacemakers			A26	Insufficient Device Problem Information		
3854	2015				CT257	Implantable cardiac pacemakers			A0203	Defective Device		
									A07	Electrical /Electronic Property Problem		
									A0709	Device Sensing Problem		
3855	2015		47270	Cardiac resynchronization therapy implantable defibrillator	CT2866	Implantable cardiac electrical stimulators			A01	Patient Device Interaction Problem		
									A17	Compatibility Problem		
3856	2016	TBD	35223	Endocardial/interventricular septal pacing lead	CT1684	Cardiac pacemakers and associated devices			A26	Insufficient Device Problem Information		
3857	2015	All other reportable incidents			CT257	Implantable cardiac pacemakers			A07	Electrical /Electronic Property Problem		
									A0708	Power Problem		

3858	2016	All other reportable incidents	47270	Cardiac resynchronization therapy implantable defibrillator	CT2866	Implantable cardiac electrical stimulators			A07	Electrical /Electronic Property Problem	14/10/2016	
									A0708	Power Problem	14/10/2016	
3859	2015		47270	Cardiac resynchronization therapy implantable defibrillator	CT2866	Implantable cardiac electrical stimulators			A07	Electrical /Electronic Property Problem		
									A0709	Device Sensing Problem		
3860	2016	TBD			CT257	Implantable cardiac pacemakers			A07	Electrical /Electronic Property Problem		
									A0708	Power Problem		
3861	2015		47270	Cardiac resynchronization therapy implantable defibrillator	CT2866	Implantable cardiac electrical stimulators			A26	Insufficient Device Problem Information		
3862	2015		47270	Cardiac resynchronization therapy implantable defibrillator	CT2866	Implantable cardiac electrical stimulators			A26	Insufficient Device Problem Information		
3863	2016	All other reportable incidents			CT257	Implantable cardiac pacemakers			A07	Electrical /Electronic Property Problem	06/04/2016	
									A0708	Power Problem	06/04/2016	
3864	2016	TBD			CT257	Implantable cardiac pacemakers			A26	Insufficient Device Problem Information		
3865	2016	All other reportable incidents	47270	Cardiac resynchronization therapy implantable defibrillator	CT2866	Implantable cardiac electrical stimulators			A07	Electrical /Electronic Property Problem	10/10/2016	
									A0708	Power Problem	10/10/2016	
3866	2015	All other reportable incidents			CT257	Implantable cardiac pacemakers			A07	Electrical /Electronic Property Problem		
									A0708	Power Problem		
3867	2015	All other reportable incidents	47270	Cardiac resynchronization therapy implantable defibrillator	CT2866	Implantable cardiac electrical stimulators			A07	Electrical /Electronic Property Problem		
									A0708	Power Problem		
3868	2015				CT257	Implantable cardiac pacemakers			A07	Electrical /Electronic Property Problem		
									A0708	Power Problem		
3869	2015	All other reportable incidents			CT257	Implantable cardiac pacemakers			A07	Electrical /Electronic Property Problem		
									A0708	Power Problem		
3870	2015	All other reportable incidents			CT257	Implantable cardiac pacemakers			A07	Electrical /Electronic Property Problem		
									A0708	Power Problem		
3871	2016	TBD	47270	Cardiac resynchronization therapy implantable defibrillator	CT2866	Implantable cardiac electrical stimulators			A07	Electrical /Electronic Property Problem		
									A0708	Power Problem		
3872	2016	TBD	47270	Cardiac resynchronization therapy implantable defibrillator	CT2866	Implantable cardiac electrical stimulators			A07	Electrical /Electronic Property Problem		
									A0709	Device Sensing Problem		
3873	2016	All other reportable incidents			CT257	Implantable cardiac pacemakers			A07	Electrical /Electronic Property Problem	17/05/2016	
									A0708	Power Problem	17/05/2016	
3874	2016	All other reportable incidents	47270	Cardiac resynchronization therapy implantable defibrillator	CT2866	Implantable cardiac electrical stimulators			A07	Electrical /Electronic Property Problem	24/10/2016	
									A0708	Power Problem	24/10/2016	
3875	2016	All other reportable incidents	47270	Cardiac resynchronization therapy implantable defibrillator	CT2866	Implantable cardiac electrical stimulators			A07	Electrical /Electronic Property Problem	25/10/2016	
									A0708	Power Problem	25/10/2016	
3876	2015		47270	Cardiac resynchronization therapy implantable defibrillator	CT2866	Implantable cardiac electrical stimulators			A11	Computer Software Problem		
									A1102	Application Program Problem		
3877	2015				CT257	Implantable cardiac pacemakers			A07	Electrical /Electronic Property Problem		
									A0708	Power Problem		
3878	2015				CT257	Implantable cardiac pacemakers			A13	Communication or Transmission Problem		
									A26	Insufficient Device Problem Information		
3879	2016	TBD			CT257	Implantable cardiac pacemakers			A07	Electrical /Electronic Property Problem	06/06/2016	
									A0708	Power Problem	06/06/2016	
3880	2015				CT257	Implantable cardiac pacemakers			A11	Computer Software Problem		
									A1102	Application Program Problem		
3881	2016	All other reportable incidents			CT257	Implantable cardiac pacemakers			A07	Electrical /Electronic Property Problem	01/10/2016	
									A0708	Power Problem	01/10/2016	
3882	2016	TBD	47270	Cardiac resynchronization therapy implantable defibrillator	CT2866	Implantable cardiac electrical stimulators			A13	Communication or Transmission Problem		
									A26	Insufficient Device Problem Information		
3883	2015				CT257	Implantable cardiac pacemakers			A07	Electrical /Electronic Property Problem		
									A0708	Power Problem		
3884	2016	All other reportable incidents	47270	Cardiac resynchronization therapy implantable defibrillator	CT2866	Implantable cardiac electrical stimulators			A07	Electrical /Electronic Property Problem		
									A0708	Power Problem		
3885	2016	TBD	35852	Single-chamber implantable defibrillator	CT2866	Implantable cardiac electrical stimulators			A01	Patient Device Interaction Problem		
									A17	Compatibility Problem		
3886	2016	All other reportable incidents			CT257	Implantable cardiac pacemakers			A07	Electrical /Electronic Property Problem	10/10/2016	
									A0708	Power Problem	10/10/2016	
3887	2015				CT257	Implantable cardiac pacemakers			A26	Insufficient Device Problem Information		
3888	2016	All other reportable incidents			CT257	Implantable cardiac pacemakers			A07	Electrical /Electronic Property Problem	27/05/2016	
									A0708	Power Problem	27/05/2016	
3889	2016	TBD			CT257	Implantable cardiac pacemakers			A09	Output Problem		
									A0907	No Device Output		
3890	2015				CT257	Implantable cardiac pacemakers			A07	Electrical /Electronic Property Problem		
									A0708	Power Problem		
3891	2016	All other reportable incidents	47270	Cardiac resynchronization therapy implantable defibrillator	CT2866	Implantable cardiac electrical stimulators			A07	Electrical /Electronic Property Problem	16/12/2015	
									A0708	Power Problem	16/12/2015	
3892	2015				CT257	Implantable cardiac pacemakers			A09	Output Problem		
									A090404	Output Above Specifications		
									A16	Protective Measures Problem		
									A1602	Fail-Safe Problem		
3893	2015		47270	Cardiac resynchronization therapy implantable defibrillator	CT2866	Implantable cardiac electrical stimulators			A26	Insufficient Device Problem Information		
3894	2015		61438	Cardiac pulse generator reader, home-use	CT1684	Cardiac pacemakers and associated devices			A07	Electrical /Electronic Property Problem		
									A0708	Power Problem		
3895	2016	All other reportable incidents			CT257	Implantable cardiac pacemakers			A07	Electrical /Electronic Property Problem	13/10/2016	
									A0708	Power Problem	13/10/2016	
3896	2016	All other reportable incidents	47270	Cardiac resynchronization therapy implantable defibrillator	CT2866	Implantable cardiac electrical stimulators			A07	Electrical /Electronic Property Problem	10/10/2016	
									A0708	Power Problem	10/10/2016	
3897	2016	All other reportable incidents			CT257	Implantable cardiac pacemakers			A07	Electrical /Electronic Property Problem		
									A0708	Power Problem		
3898	2016	TBD	47270	Cardiac resynchronization therapy implantable defibrillator	CT2866	Implantable cardiac electrical stimulators			A26	Insufficient Device Problem Information		
3899	2015				CT257	Implantable cardiac pacemakers			A26	Insufficient Device Problem Information		
3900	2016	All other reportable incidents			CT257	Implantable cardiac pacemakers			A07	Electrical /Electronic Property Problem		
									A0708	Power Problem		
3901	2016	TBD	35223	Endocardial/interventricular septal pacing lead	CT1684	Cardiac pacemakers and associated devices			A26	Insufficient Device Problem Information		
3902	2016	All other reportable incidents			CT257	Implantable cardiac pacemakers			A1304	Telemetry Discrepancy		
									A26	Insufficient Device Problem Information		

3903	2016	TBD	35223	Endocardial/interventricular septal pacing lead	CT1684	Cardiac pacemakers and associated devices			A26	Insufficient Device Problem Information		
3904	2016	TBD			CT257	Implantable cardiac pacemakers			A07	Electrical /Electronic Property Problem		
									A0708	Power Problem		
3905	2016	TBD	35223	Endocardial/interventricular septal pacing lead	CT1684	Cardiac pacemakers and associated devices			A15	Activation, Positioning or Separation Problem		
									A150203	Difficult or Delayed Positioning		
3906	2016	All other reportable incidents			CT257	Implantable cardiac pacemakers			A07	Electrical /Electronic Property Problem		
									A0708	Power Problem		
3907	2015		47270	Cardiac resynchronization therapy implantable defibrillator	CT2866	Implantable cardiac electrical stimulators			A07	Electrical /Electronic Property Problem		
									A0708	Power Problem		
3908	2015		47270	Cardiac resynchronization therapy implantable defibrillator	CT2866	Implantable cardiac electrical stimulators			A26	Insufficient Device Problem Information		
3909	2015				CT257	Implantable cardiac pacemakers			A07	Electrical /Electronic Property Problem		
									A0708	Power Problem		
3910	2015				CT257	Implantable cardiac pacemakers			A07	Electrical /Electronic Property Problem		
									A0708	Power Problem		
3911	2015				CT1684	Cardiac pacemakers and associated devices			A07	Electrical /Electronic Property Problem		
									A0708	Power Problem		
3912	2016	All other reportable incidents	47263	Cardiac resynchronization therapy implantable pacemaker	CT1684	Cardiac pacemakers and associated devices			A07	Electrical /Electronic Property Problem	29/12/2015	
									A0708	Power Problem	29/12/2015	
					CT1685	Cardiac pacemakers			A07	Electrical /Electronic Property Problem	29/12/2015	
									A0708	Power Problem	29/12/2015	
					CT257	Implantable cardiac pacemakers			A07	Electrical /Electronic Property Problem	29/12/2015	
									A0708	Power Problem	29/12/2015	
					CT2866	Implantable cardiac electrical stimulators			A07	Electrical /Electronic Property Problem	29/12/2015	
									A0708	Power Problem	29/12/2015	
3913	2015		35852	Single-chamber implantable defibrillator	CT2866	Implantable cardiac electrical stimulators			A01	Patient Device Interaction Problem		
									A17	Compatibility Problem		
3914	2016	All other reportable incidents			CT257	Implantable cardiac pacemakers			A07	Electrical /Electronic Property Problem	03/12/2015	
									A0708	Power Problem	03/12/2015	
3915	2016	All other reportable incidents			CT257	Implantable cardiac pacemakers			A07	Electrical /Electronic Property Problem	06/01/2016	
									A0708	Power Problem	06/01/2016	
3916	2015		47270	Cardiac resynchronization therapy implantable defibrillator	CT2866	Implantable cardiac electrical stimulators			A26	Insufficient Device Problem Information		
3917	2016	Death	47270	Cardiac resynchronization therapy implantable defibrillator	CT2866	Implantable cardiac electrical stimulators			A26	Insufficient Device Problem Information		
3918	2016	TBD			CT257	Implantable cardiac pacemakers			A26	Insufficient Device Problem Information		
3919	2016	TBD			CT1684	Cardiac pacemakers and associated devices			A09	Output Problem		
									A0907	No Device Output		
3920	2015				CT257	Implantable cardiac pacemakers			A26	Insufficient Device Problem Information		
3921	2016	TBD			CT1684	Cardiac pacemakers and associated devices			A090804	False Positive Result		
									A26	Insufficient Device Problem Information		
3922	2016	TBD			CT257	Implantable cardiac pacemakers			A07	Electrical /Electronic Property Problem		
									A0708	Power Problem		
3923	2016	All other reportable incidents	47263	Cardiac resynchronization therapy implantable pacemaker	CT1684	Cardiac pacemakers and associated devices			A07	Electrical /Electronic Property Problem	05/04/2016	
									A0708	Power Problem	05/04/2016	
					CT1685	Cardiac pacemakers			A07	Electrical /Electronic Property Problem	05/04/2016	
									A0708	Power Problem	05/04/2016	
					CT257	Implantable cardiac pacemakers			A07	Electrical /Electronic Property Problem	05/04/2016	
									A0708	Power Problem	05/04/2016	
					CT2866	Implantable cardiac electrical stimulators			A07	Electrical /Electronic Property Problem	05/04/2016	
									A0708	Power Problem	05/04/2016	
3924	2016	All other reportable incidents	47270	Cardiac resynchronization therapy implantable defibrillator	CT2866	Implantable cardiac electrical stimulators			A07	Electrical /Electronic Property Problem	14/10/2016	
									A0708	Power Problem	14/10/2016	
3925	2016	All other reportable incidents			CT257	Implantable cardiac pacemakers			A07	Electrical /Electronic Property Problem	01/11/2016	
									A0708	Power Problem	01/11/2016	
									A13	Communication or Transmission Problem	01/11/2016	
									A26	Insufficient Device Problem Information	01/11/2016	
3926	2016	All other reportable incidents	47270	Cardiac resynchronization therapy implantable defibrillator	CT2866	Implantable cardiac electrical stimulators			A07	Electrical /Electronic Property Problem	19/10/2016	
									A0708	Power Problem	19/10/2016	
3927	2015				CT1684	Cardiac pacemakers and associated devices			A07	Electrical /Electronic Property Problem		
									A0709	Device Sensing Problem		
									A09	Output Problem		
									A0907	No Device Output		
3928	2015				CT257	Implantable cardiac pacemakers			A16	Protective Measures Problem		
									A1602	Fail-Safe Problem		
3929	2015				CT257	Implantable cardiac pacemakers			A26	Insufficient Device Problem Information		
3930	2015		47270	Cardiac resynchronization therapy implantable defibrillator	CT2866	Implantable cardiac electrical stimulators			A26	Insufficient Device Problem Information		
3931	2016	TBD	47263	Cardiac resynchronization therapy implantable pacemaker	CT1684	Cardiac pacemakers and associated devices			A07	Electrical /Electronic Property Problem		
									A0708	Power Problem		
					CT1685	Cardiac pacemakers			A07	Electrical /Electronic Property Problem		
									A0708	Power Problem		
					CT257	Implantable cardiac pacemakers			A07	Electrical /Electronic Property Problem		
									A0708	Power Problem		
					CT2866	Implantable cardiac electrical stimulators			A07	Electrical /Electronic Property Problem		
									A0708	Power Problem		
3932	2016	TBD	47270	Cardiac resynchronization therapy implantable defibrillator	CT2866	Implantable cardiac electrical stimulators			A07	Electrical /Electronic Property Problem	12/11/2016	
									A0708	Power Problem	12/11/2016	
3933	2015				CT257	Implantable cardiac pacemakers			A1304	Telemetry Discrepancy		
									A26	Insufficient Device Problem Information		
3934	2016	All other reportable incidents	47270	Cardiac resynchronization therapy implantable defibrillator	CT2866	Implantable cardiac electrical stimulators			A07	Electrical /Electronic Property Problem	03/11/2016	
									A0708	Power Problem	03/11/2016	
3935	2016	TBD	35223	Endocardial/interventricular septal pacing lead	CT1684	Cardiac pacemakers and associated devices			A26	Insufficient Device Problem Information		
3936	2015				CT257	Implantable cardiac pacemakers			A26	Insufficient Device Problem Information		
3937	2016	TBD	47270	Cardiac resynchronization therapy implantable defibrillator	CT2866	Implantable cardiac electrical stimulators			A09	Output Problem		
									A090404	Output Above Specifications		
3938	2016	All other reportable incidents	35223	Endocardial/interventricular septal pacing lead	CT1684	Cardiac pacemakers and associated devices			A07	Electrical /Electronic Property Problem		

									A0709	Device Sensing Problem		
3939	2016	TBD	35223	Endocardial/interventricular septal pacing lead	CT1684	Cardiac pacemakers and associated devices			A26	Insufficient Device Problem Information		
3940	2016	All other reportable incidents	47270	Cardiac resynchronization therapy implantable defibrillator	CT2866	Implantable cardiac electrical stimulators			A07	Electrical /Electronic Property Problem	31/12/2015	
									A0708	Power Problem	31/12/2015	
3941	2016	TBD	47270	Cardiac resynchronization therapy implantable defibrillator	CT2866	Implantable cardiac electrical stimulators			A01	Patient Device Interaction Problem	04/11/2016	
									A17	Compatibility Problem	04/11/2016	
3942	2015		35852	Single-chamber implantable defibrillator	CT2866	Implantable cardiac electrical stimulators			A07	Electrical /Electronic Property Problem	27/08/2015	
									A0708	Power Problem	27/08/2015	
3943	2016	All other reportable incidents	47270	Cardiac resynchronization therapy implantable defibrillator	CT2866	Implantable cardiac electrical stimulators			A07	Electrical /Electronic Property Problem		
									A0708	Power Problem		
3944	2016	Death			CT257	Implantable cardiac pacemakers			A26	Insufficient Device Problem Information		
3945	2016	TBD			CT257	Implantable cardiac pacemakers			A07	Electrical /Electronic Property Problem		
									A0708	Power Problem		
3946	2016	All other reportable incidents	47270	Cardiac resynchronization therapy implantable defibrillator	CT2866	Implantable cardiac electrical stimulators			A07	Electrical /Electronic Property Problem	16/10/2016	
									A0708	Power Problem	16/10/2016	
3947	2015		47270	Cardiac resynchronization therapy implantable defibrillator	CT2866	Implantable cardiac electrical stimulators			A07	Electrical /Electronic Property Problem		
									A0708	Power Problem		
3948	2015				CT257	Implantable cardiac pacemakers			A07	Electrical /Electronic Property Problem		
									A0708	Power Problem		
3949	2016	TBD			CT257	Implantable cardiac pacemakers			A13	Communication or Transmission Problem		
									A26	Insufficient Device Problem Information		
3950	2016	TBD	47270	Cardiac resynchronization therapy implantable defibrillator	CT2866	Implantable cardiac electrical stimulators			A07	Electrical /Electronic Property Problem		
									A0708	Power Problem		
3951	2016	All other reportable incidents	47263	Cardiac resynchronization therapy implantable pacemaker	CT1684	Cardiac pacemakers and associated devices			A07	Electrical /Electronic Property Problem	26/09/2016	
									A0708	Power Problem	26/09/2016	
					CT1685	Cardiac pacemakers			A07	Electrical /Electronic Property Problem	26/09/2016	
									A0708	Power Problem	26/09/2016	
					CT257	Implantable cardiac pacemakers			A07	Electrical /Electronic Property Problem	26/09/2016	
									A0708	Power Problem	26/09/2016	
					CT2866	Implantable cardiac electrical stimulators			A07	Electrical /Electronic Property Problem	26/09/2016	
									A0708	Power Problem	26/09/2016	
3952	2016	TBD	47270	Cardiac resynchronization therapy implantable defibrillator	CT2866	Implantable cardiac electrical stimulators			A07	Electrical /Electronic Property Problem		
									A0708	Power Problem		
3953	2015				CT257	Implantable cardiac pacemakers			A07	Electrical /Electronic Property Problem	27/08/2015	
									A0708	Power Problem	27/08/2015	
3954	2016	TBD	47270	Cardiac resynchronization therapy implantable defibrillator	CT2866	Implantable cardiac electrical stimulators			A07	Electrical /Electronic Property Problem		
									A0708	Power Problem		
3955	2015				CT257	Implantable cardiac pacemakers			A07	Electrical /Electronic Property Problem		
									A0708	Power Problem		
3956	2016	TBD	35223	Endocardial/interventricular septal pacing lead	CT1684	Cardiac pacemakers and associated devices			A26	Insufficient Device Problem Information		
3957	2015				CT257	Implantable cardiac pacemakers			A17	Compatibility Problem		
									A1702	Device-Device Incompatibility		
3958	2016	TBD			CT257	Implantable cardiac pacemakers			A13	Communication or Transmission Problem		
									A26	Insufficient Device Problem Information		
3959	2015	All other reportable incidents	47270	Cardiac resynchronization therapy implantable defibrillator	CT2866	Implantable cardiac electrical stimulators			A1304	Telemetry Discrepancy	18/08/2015	
									A26	Insufficient Device Problem Information	18/08/2015	
3960	2015				CT257	Implantable cardiac pacemakers			A26	Insufficient Device Problem Information		
3961	2016	TBD	47270	Cardiac resynchronization therapy implantable defibrillator	CT2866	Implantable cardiac electrical stimulators			A07	Electrical /Electronic Property Problem		
									A0708	Power Problem		
3962	2016	All other reportable incidents			CT257	Implantable cardiac pacemakers			A16	Protective Measures Problem		
									A1602	Fail-Safe Problem		
3963	2015		47270	Cardiac resynchronization therapy implantable defibrillator	CT2866	Implantable cardiac electrical stimulators			A26	Insufficient Device Problem Information		
3964	2016	TBD			CT257	Implantable cardiac pacemakers			A26	Insufficient Device Problem Information		
3965	2016	TBD			CT257	Implantable cardiac pacemakers			A13	Communication or Transmission Problem		
									A26	Insufficient Device Problem Information		
3966	2016	All other reportable incidents	47270	Cardiac resynchronization therapy implantable defibrillator	CT2866	Implantable cardiac electrical stimulators			A07	Electrical /Electronic Property Problem		
									A0709	Device Sensing Problem		
3967	2016	TBD			CT1684	Cardiac pacemakers and associated devices			A09	Output Problem		
									A090404	Output Above Specifications		
3968	2016	TBD	47270	Cardiac resynchronization therapy implantable defibrillator	CT2866	Implantable cardiac electrical stimulators			A26	Insufficient Device Problem Information		
3969	2016	TBD	47270	Cardiac resynchronization therapy implantable defibrillator	CT2866	Implantable cardiac electrical stimulators			A07	Electrical /Electronic Property Problem		
									A0709	Device Sensing Problem		
3970	2015				CT257	Implantable cardiac pacemakers			A07	Electrical /Electronic Property Problem		
									A0708	Power Problem		
3971	2016	All other reportable incidents			CT257	Implantable cardiac pacemakers			A16	Protective Measures Problem		
									A1602	Fail-Safe Problem		
3972	2016	TBD	47270	Cardiac resynchronization therapy implantable defibrillator	CT2866	Implantable cardiac electrical stimulators			A07	Electrical /Electronic Property Problem	14/11/2016	
									A0708	Power Problem	14/11/2016	
3973	2016	All other reportable incidents	47270	Cardiac resynchronization therapy implantable defibrillator	CT2866	Implantable cardiac electrical stimulators			A13	Communication or Transmission Problem		
									A26	Insufficient Device Problem Information		
3974	2016	All other reportable incidents			CT257	Implantable cardiac pacemakers			A07	Electrical /Electronic Property Problem	25/10/2016	
									A0708	Power Problem	25/10/2016	
3975	2015		47270	Cardiac resynchronization therapy implantable defibrillator	CT2866	Implantable cardiac electrical stimulators			A07	Electrical /Electronic Property Problem		
									A0709	Device Sensing Problem		
3976	2016	All other reportable incidents			CT1684	Cardiac pacemakers and associated devices			A07	Electrical /Electronic Property Problem		
									A0708	Power Problem		
									A09	Output Problem		
									A0907	No Device Output		
3977	2016	All other reportable incidents			CT1684	Cardiac pacemakers and associated devices			A09	Output Problem		
									A0907	No Device Output		
3978	2016	TBD			CT257	Implantable cardiac pacemakers			A13	Communication or Transmission Problem		
									A26	Insufficient Device Problem Information		
3979	2016	TBD			CT257	Implantable cardiac pacemakers			A07	Electrical /Electronic Property Problem		
									A0708	Power Problem		

3980	2015		47270	Cardiac resynchronization therapy implantable defibrillator	CT2866	Implantable cardiac electrical stimulators			A26	Insufficient Device Problem Information		
3981	2016	TBD			CT1684	Cardiac pacemakers and associated devices			A07	Electrical /Electronic Property Problem		
									A0708	Power Problem		
3982	2015		35852	Single-chamber implantable defibrillator	CT2866	Implantable cardiac electrical stimulators			A01	Patient Device Interaction Problem		
									A17	Compatibility Problem		
3983	2015		47270	Cardiac resynchronization therapy implantable defibrillator	CT2866	Implantable cardiac electrical stimulators			A26	Insufficient Device Problem Information		
3984	2015				CT257	Implantable cardiac pacemakers			A26	Insufficient Device Problem Information		
3985	2016	TBD			CT257	Implantable cardiac pacemakers			A07	Electrical /Electronic Property Problem	20/01/2016	
									A0708	Power Problem	20/01/2016	
3986	2015		47270	Cardiac resynchronization therapy implantable defibrillator	CT2866	Implantable cardiac electrical stimulators			A07	Electrical /Electronic Property Problem		
									A0709	Device Sensing Problem		
3987	2015		47270	Cardiac resynchronization therapy implantable defibrillator	CT2866	Implantable cardiac electrical stimulators			A07	Electrical /Electronic Property Problem		
									A0708	Power Problem		
3988	2016	TBD	47270	Cardiac resynchronization therapy implantable defibrillator	CT2866	Implantable cardiac electrical stimulators			A07	Electrical /Electronic Property Problem		
									A0708	Power Problem		
3989	2016	TBD	35223	Endocardial/interventricular septal pacing lead	CT1684	Cardiac pacemakers and associated devices			A26	Insufficient Device Problem Information		
3990	2015	All other reportable incidents	47270	Cardiac resynchronization therapy implantable defibrillator	CT2866	Implantable cardiac electrical stimulators			A07	Electrical /Electronic Property Problem	17/08/2015	
									A0708	Power Problem	17/08/2015	
3991	2016	All other reportable incidents	47263	Cardiac resynchronization therapy implantable pacemaker	CT1684	Cardiac pacemakers and associated devices			A07	Electrical /Electronic Property Problem	03/12/2015	
									A0708	Power Problem	03/12/2015	
					CT1685	Cardiac pacemakers			A07	Electrical /Electronic Property Problem	03/12/2015	
									A0708	Power Problem	03/12/2015	
					CT257	Implantable cardiac pacemakers			A07	Electrical /Electronic Property Problem	03/12/2015	
									A0708	Power Problem	03/12/2015	
					CT2866	Implantable cardiac electrical stimulators			A07	Electrical /Electronic Property Problem	03/12/2015	
									A0708	Power Problem	03/12/2015	
3992	2016	All other reportable incidents	47263	Cardiac resynchronization therapy implantable pacemaker	CT1684	Cardiac pacemakers and associated devices			A07	Electrical /Electronic Property Problem	09/12/2015	
									A0708	Power Problem	09/12/2015	
					CT1685	Cardiac pacemakers			A07	Electrical /Electronic Property Problem	09/12/2015	
									A0708	Power Problem	09/12/2015	
					CT257	Implantable cardiac pacemakers			A07	Electrical /Electronic Property Problem	09/12/2015	
									A0708	Power Problem	09/12/2015	
					CT2866	Implantable cardiac electrical stimulators			A07	Electrical /Electronic Property Problem	09/12/2015	
									A0708	Power Problem	09/12/2015	
3993	2016	All other reportable incidents	47263	Cardiac resynchronization therapy implantable pacemaker	CT1684	Cardiac pacemakers and associated devices			A07	Electrical /Electronic Property Problem	07/12/2015	
									A0708	Power Problem	07/12/2015	
					CT1685	Cardiac pacemakers			A07	Electrical /Electronic Property Problem	07/12/2015	
									A0708	Power Problem	07/12/2015	
					CT257	Implantable cardiac pacemakers			A07	Electrical /Electronic Property Problem	07/12/2015	
									A0708	Power Problem	07/12/2015	
					CT2866	Implantable cardiac electrical stimulators			A07	Electrical /Electronic Property Problem	07/12/2015	
									A0708	Power Problem	07/12/2015	
3994	2015	All other reportable incidents			CT257	Implantable cardiac pacemakers			A07	Electrical /Electronic Property Problem	30/03/2015	
									A0708	Power Problem	30/03/2015	
3995	2016	TBD	47270	Cardiac resynchronization therapy implantable defibrillator	CT2866	Implantable cardiac electrical stimulators			A26	Insufficient Device Problem Information		
3996	2015		47270	Cardiac resynchronization therapy implantable defibrillator	CT2866	Implantable cardiac electrical stimulators			A07	Electrical /Electronic Property Problem		
									A0708	Power Problem		
3997	2015		47270	Cardiac resynchronization therapy implantable defibrillator	CT2866	Implantable cardiac electrical stimulators			A07	Electrical /Electronic Property Problem	08/07/2015	
									A0708	Power Problem	08/07/2015	
									A26	Insufficient Device Problem Information	08/07/2015	
3998	2016	TBD	47270	Cardiac resynchronization therapy implantable defibrillator	CT2866	Implantable cardiac electrical stimulators			A26	Insufficient Device Problem Information		
3999	2016	TBD	35223	Endocardial/interventricular septal pacing lead	CT1684	Cardiac pacemakers and associated devices			A26	Insufficient Device Problem Information		
4000	2015		35852	Single-chamber implantable defibrillator	CT2866	Implantable cardiac electrical stimulators			A07	Electrical /Electronic Property Problem		
									A0709	Device Sensing Problem		
									A09	Output Problem		
									A090404	Output Above Specifications		
4001	2015				CT257	Implantable cardiac pacemakers			A07	Electrical /Electronic Property Problem		
									A0708	Power Problem		
4002	2015				CT257	Implantable cardiac pacemakers			A07	Electrical /Electronic Property Problem		
									A0708	Power Problem		
									A16	Protective Measures Problem		
									A1602	Fail-Safe Problem		
4003	2015				CT257	Implantable cardiac pacemakers			A26	Insufficient Device Problem Information		
4004	2016	TBD			CT257	Implantable cardiac pacemakers			A26	Insufficient Device Problem Information		
4005	2015				CT257	Implantable cardiac pacemakers			A13	Communication or Transmission Problem		
									A26	Insufficient Device Problem Information		
4006	2015				CT257	Implantable cardiac pacemakers			A26	Insufficient Device Problem Information		
4007	2016	TBD	47270	Cardiac resynchronization therapy implantable defibrillator	CT2866	Implantable cardiac electrical stimulators			A07	Electrical /Electronic Property Problem		
									A0708	Power Problem		
4008	2015		35223	Endocardial/interventricular septal pacing lead	CT1684	Cardiac pacemakers and associated devices			A07	Electrical /Electronic Property Problem		
									A0709	Device Sensing Problem		
4009	2016	All other reportable incidents	47270	Cardiac resynchronization therapy implantable defibrillator	CT2866	Implantable cardiac electrical stimulators			A16	Protective Measures Problem	06/06/2016	
									A1602	Fail-Safe Problem	06/06/2016	
4010	2016	All other reportable incidents			CT257	Implantable cardiac pacemakers			A09	Output Problem	01/06/2016	
									A0907	No Device Output	01/06/2016	
4011	2016	TBD			CT257	Implantable cardiac pacemakers			A07	Electrical /Electronic Property Problem		
									A0708	Power Problem		
4012	2016	TBD	35223	Endocardial/interventricular septal pacing lead	CT1684	Cardiac pacemakers and associated devices			A26	Insufficient Device Problem Information		
4013	2015				CT257	Implantable cardiac pacemakers			A11	Computer Software Problem		
									A1102	Application Program Problem		
4014	2016	All other reportable incidents	47270	Cardiac resynchronization therapy implantable defibrillator	CT2866	Implantable cardiac electrical stimulators			A13	Communication or Transmission Problem		
									A26	Insufficient Device Problem Information		

4015	2016	TBD	47270	Cardiac resynchronization therapy implantable defibrillator	CT2866	Implantable cardiac electrical stimulators			A07	Electrical /Electronic Property Problem		
									A0709	Device Sensing Problem		
4016	2015	All other reportable incidents			CT257	Implantable cardiac pacemakers			A07	Electrical /Electronic Property Problem		
									A0708	Power Problem		
4017	2015				CT1684	Cardiac pacemakers and associated devices			A09	Output Problem		
									A0907	No Device Output		
4018	2016	All other reportable incidents	47270	Cardiac resynchronization therapy implantable defibrillator	CT2866	Implantable cardiac electrical stimulators			A13	Communication or Transmission Problem		
									A26	Insufficient Device Problem Information		
4019	2015				CT257	Implantable cardiac pacemakers			A09	Output Problem		
									A090404	Output Above Specifications		
									A11	Computer Software Problem		
									A1102	Application Program Problem		
4020	2015	All other reportable incidents			CT257	Implantable cardiac pacemakers			A07	Electrical /Electronic Property Problem	02/04/2015	
									A0708	Power Problem	02/04/2015	
4021	2015	All other reportable incidents			CT257	Implantable cardiac pacemakers			A07	Electrical /Electronic Property Problem	02/07/2015	
									A0708	Power Problem	02/07/2015	
4022	2015	All other reportable incidents	47270	Cardiac resynchronization therapy implantable defibrillator	CT2866	Implantable cardiac electrical stimulators			A07	Electrical /Electronic Property Problem		
									A0708	Power Problem		
4023	2016	TBD			CT257	Implantable cardiac pacemakers			A07	Electrical /Electronic Property Problem		
									A0708	Power Problem		
4024	2015	All other reportable incidents	35223	Endocardial/interventricular septal pacing lead	CT1684	Cardiac pacemakers and associated devices			A04	Material Integrity Problem		
									A0405	Degraded		
4025	2016	TBD	35223	Endocardial/interventricular septal pacing lead	CT1684	Cardiac pacemakers and associated devices			A26	Insufficient Device Problem Information		
4026	2015	All other reportable incidents			CT257	Implantable cardiac pacemakers			A07	Electrical /Electronic Property Problem		
									A0708	Power Problem		
4027	2015		47270	Cardiac resynchronization therapy implantable defibrillator	CT2866	Implantable cardiac electrical stimulators			A26	Insufficient Device Problem Information		
4028	2015				CT257	Implantable cardiac pacemakers			A26	Insufficient Device Problem Information		
4029	2015		47270	Cardiac resynchronization therapy implantable defibrillator	CT2866	Implantable cardiac electrical stimulators			A26	Insufficient Device Problem Information		
4030	2015		47270	Cardiac resynchronization therapy implantable defibrillator	CT2866	Implantable cardiac electrical stimulators			A17	Compatibility Problem		
									A1702	Device-Device Incompatibility		
4031	2015				CT257	Implantable cardiac pacemakers			A07	Electrical /Electronic Property Problem		
									A0708	Power Problem		
4032	2016	All other reportable incidents	47270	Cardiac resynchronization therapy implantable defibrillator	CT2866	Implantable cardiac electrical stimulators			A07	Electrical /Electronic Property Problem	17/10/2016	
									A0708	Power Problem	17/10/2016	
4033	2016	All other reportable incidents			CT257	Implantable cardiac pacemakers			A07	Electrical /Electronic Property Problem	21/01/2016	
									A0708	Power Problem	21/01/2016	
4034	2016	All other reportable incidents			CT257	Implantable cardiac pacemakers			A07	Electrical /Electronic Property Problem	16/11/2015	
									A0708	Power Problem	16/11/2015	
4035	2016	TBD	47270	Cardiac resynchronization therapy implantable defibrillator	CT2866	Implantable cardiac electrical stimulators			A07	Electrical /Electronic Property Problem		
									A0708	Power Problem		
4036	2015		47270	Cardiac resynchronization therapy implantable defibrillator	CT2866	Implantable cardiac electrical stimulators			A26	Insufficient Device Problem Information		
4037	2016	All other reportable incidents			CT257	Implantable cardiac pacemakers			A07	Electrical /Electronic Property Problem	09/11/2016	
									A0709	Device Sensing Problem	09/11/2016	
4038	2015		47270	Cardiac resynchronization therapy implantable defibrillator	CT2866	Implantable cardiac electrical stimulators			A26	Insufficient Device Problem Information		
4039	2015	All other reportable incidents			CT257	Implantable cardiac pacemakers			A07	Electrical /Electronic Property Problem		
									A0708	Power Problem		
4040	2016	TBD			CT257	Implantable cardiac pacemakers			A26	Insufficient Device Problem Information		
4041	2015		47270	Cardiac resynchronization therapy implantable defibrillator	CT2866	Implantable cardiac electrical stimulators			A26	Insufficient Device Problem Information		
4042	2016	Death	47270	Cardiac resynchronization therapy implantable defibrillator	CT2866	Implantable cardiac electrical stimulators			A24	No Apparent Device Problem		
									A26	Insufficient Device Problem Information		
4043	2016	All other reportable incidents	47270	Cardiac resynchronization therapy implantable defibrillator	CT2866	Implantable cardiac electrical stimulators			A07	Electrical /Electronic Property Problem	10/03/2016	
									A0708	Power Problem	10/03/2016	
4044	2015	All other reportable incidents	47270	Cardiac resynchronization therapy implantable defibrillator	CT2866	Implantable cardiac electrical stimulators			A07	Electrical /Electronic Property Problem	08/09/2015	
									A0708	Power Problem	08/09/2015	
4045	2016	TBD	35223	Endocardial/interventricular septal pacing lead	CT1684	Cardiac pacemakers and associated devices			A26	Insufficient Device Problem Information		
4046	2015				CT257	Implantable cardiac pacemakers			A07	Electrical /Electronic Property Problem		
									A0708	Power Problem		
4047	2015				CT257	Implantable cardiac pacemakers			A26	Insufficient Device Problem Information		
4048	2016	TBD			CT257	Implantable cardiac pacemakers			A26	Insufficient Device Problem Information		
4049	2016	TBD	47270	Cardiac resynchronization therapy implantable defibrillator	CT2866	Implantable cardiac electrical stimulators			A07	Electrical /Electronic Property Problem		
									A0709	Device Sensing Problem		
4050	2015				CT257	Implantable cardiac pacemakers			A07	Electrical /Electronic Property Problem	18/09/2015	
									A0709	Device Sensing Problem	18/09/2015	
4051	2015		47270	Cardiac resynchronization therapy implantable defibrillator	CT2866	Implantable cardiac electrical stimulators			A26	Insufficient Device Problem Information		
4052	2015		47270	Cardiac resynchronization therapy implantable defibrillator	CT2866	Implantable cardiac electrical stimulators			A26	Insufficient Device Problem Information		
4053	2016	Death	47270	Cardiac resynchronization therapy implantable defibrillator	CT2866	Implantable cardiac electrical stimulators			A09	Output Problem		
									A090404	Output Above Specifications		
4054	2016	TBD			CT1684	Cardiac pacemakers and associated devices			A07	Electrical /Electronic Property Problem		
									A0708	Power Problem		
4055	2016	TBD	35223	Endocardial/interventricular septal pacing lead	CT1684	Cardiac pacemakers and associated devices			A26	Insufficient Device Problem Information		
4056	2015	All other reportable incidents			CT257	Implantable cardiac pacemakers			A07	Electrical /Electronic Property Problem	08/04/2015	
									A0708	Power Problem	08/04/2015	
4057	2016	All other reportable incidents			CT1684	Cardiac pacemakers and associated devices			A07	Electrical /Electronic Property Problem		
									A0708	Power Problem		
									A09	Output Problem		
									A0907	No Device Output		
4058	2016	All other reportable incidents			CT1684	Cardiac pacemakers and associated devices			A07	Electrical /Electronic Property Problem		
									A0708	Power Problem		
4059	2015				CT257	Implantable cardiac pacemakers			A16	Protective Measures Problem		
									A1602	Fail-Safe Problem		

4060	2016	All other reportable incidents	47263	Cardiac resynchronization therapy implantable pacemaker	CT1684	Cardiac pacemakers and associated devices			A07	Electrical /Electronic Property Problem		
									A0708	Power Problem		
					CT1685	Cardiac pacemakers			A07	Electrical /Electronic Property Problem		
									A0708	Power Problem		
					CT257	Implantable cardiac pacemakers			A07	Electrical /Electronic Property Problem		
									A0708	Power Problem		
					CT2866	Implantable cardiac electrical stimulators			A07	Electrical /Electronic Property Problem		
									A0708	Power Problem		
4061	2016	TBD	35223	Endocardial/interventricular septal pacing lead	CT1684	Cardiac pacemakers and associated devices			A15	Activation, Positioning or Separation Problem		
									A150203	Difficult or Delayed Positioning		
4062	2015				CT257	Implantable cardiac pacemakers			A26	Insufficient Device Problem Information		
4063	2015		47270	Cardiac resynchronization therapy implantable defibrillator	CT2866	Implantable cardiac electrical stimulators			A26	Insufficient Device Problem Information		
4064	2015	All other reportable incidents			CT257	Implantable cardiac pacemakers			A07	Electrical /Electronic Property Problem	07/04/2015	
									A0708	Power Problem	07/04/2015	
4065	2016	TBD	35223	Endocardial/interventricular septal pacing lead	CT1684	Cardiac pacemakers and associated devices			A26	Insufficient Device Problem Information		
4066	2016	TBD			CT257	Implantable cardiac pacemakers			A26	Insufficient Device Problem Information		
4067	2015				CT1684	Cardiac pacemakers and associated devices			A09	Output Problem		
									A090404	Output Above Specifications		
4068	2016	TBD	47270	Cardiac resynchronization therapy implantable defibrillator	CT2866	Implantable cardiac electrical stimulators			A07	Electrical /Electronic Property Problem		
									A0708	Power Problem		
4069	2016	All other reportable incidents	47263	Cardiac resynchronization therapy implantable pacemaker	CT1684	Cardiac pacemakers and associated devices			A05	Mechanical Problem		
									A051201	Device Dislodged or Dislocated		
									A09	Output Problem		
									A090404	Output Above Specifications		
					CT1685	Cardiac pacemakers			A05	Mechanical Problem		
									A051201	Device Dislodged or Dislocated		
									A09	Output Problem		
									A090404	Output Above Specifications		
					CT257	Implantable cardiac pacemakers			A05	Mechanical Problem		
									A051201	Device Dislodged or Dislocated		
									A09	Output Problem		
									A090404	Output Above Specifications		
					CT2866	Implantable cardiac electrical stimulators			A05	Mechanical Problem		
									A051201	Device Dislodged or Dislocated		
									A09	Output Problem		
									A090404	Output Above Specifications		
4070	2015		47270	Cardiac resynchronization therapy implantable defibrillator	CT2866	Implantable cardiac electrical stimulators			A07	Electrical /Electronic Property Problem		
									A0708	Power Problem		
4071	2015		47270	Cardiac resynchronization therapy implantable defibrillator	CT2866	Implantable cardiac electrical stimulators			A07	Electrical /Electronic Property Problem		
									A0708	Power Problem		
4072	2015				CT257	Implantable cardiac pacemakers			A07	Electrical /Electronic Property Problem	24/04/2015	
									A0708	Power Problem	24/04/2015	
4073	2015	All other reportable incidents			CT257	Implantable cardiac pacemakers			A07	Electrical /Electronic Property Problem		
									A0708	Power Problem		
4074	2015		47270	Cardiac resynchronization therapy implantable defibrillator	CT2866	Implantable cardiac electrical stimulators			A26	Insufficient Device Problem Information		
4075	2016	TBD	47270	Cardiac resynchronization therapy implantable defibrillator	CT2866	Implantable cardiac electrical stimulators			A07	Electrical /Electronic Property Problem		
									A0708	Power Problem		
4076	2015				CT257	Implantable cardiac pacemakers			A26	Insufficient Device Problem Information		
4077	2016	All other reportable incidents			CT257	Implantable cardiac pacemakers			A07	Electrical /Electronic Property Problem		
									A0708	Power Problem		
4078	2016	All other reportable incidents			CT257	Implantable cardiac pacemakers			A07	Electrical /Electronic Property Problem	08/11/2016	
									A0708	Power Problem	08/11/2016	
4079	2016	All other reportable incidents			CT257	Implantable cardiac pacemakers			A07	Electrical /Electronic Property Problem	20/01/2016	
									A0708	Power Problem	20/01/2016	
4080	2016	TBD	47270	Cardiac resynchronization therapy implantable defibrillator	CT2866	Implantable cardiac electrical stimulators			A07	Electrical /Electronic Property Problem		
									A0708	Power Problem		
4081	2016	All other reportable incidents	47270	Cardiac resynchronization therapy implantable defibrillator	CT2866	Implantable cardiac electrical stimulators			A07	Electrical /Electronic Property Problem	02/10/2016	
									A0708	Power Problem	02/10/2016	
4082	2016	TBD	47270	Cardiac resynchronization therapy implantable defibrillator	CT2866	Implantable cardiac electrical stimulators			A07	Electrical /Electronic Property Problem		
									A0708	Power Problem		
4083	2016	TBD	47270	Cardiac resynchronization therapy implantable defibrillator	CT2866	Implantable cardiac electrical stimulators			A07	Electrical /Electronic Property Problem		
									A0709	Device Sensing Problem		
4084	2016	TBD	47270	Cardiac resynchronization therapy implantable defibrillator	CT2866	Implantable cardiac electrical stimulators			A11	Computer Software Problem		
									A1102	Application Program Problem		
4085	2016	TBD	47270	Cardiac resynchronization therapy implantable defibrillator	CT2866	Implantable cardiac electrical stimulators			A07	Electrical /Electronic Property Problem		
									A0709	Device Sensing Problem		
4086	2016	TBD	35223	Endocardial/interventricular septal pacing lead	CT1684	Cardiac pacemakers and associated devices			A26	Insufficient Device Problem Information		
4087	2015				CT257	Implantable cardiac pacemakers			A07	Electrical /Electronic Property Problem		
									A0708	Power Problem		
4088	2015	All other reportable incidents			CT1684	Cardiac pacemakers and associated devices			A09	Output Problem		
									A0907	No Device Output		
4089	2015	All other reportable incidents			CT1684	Cardiac pacemakers and associated devices			A09	Output Problem		
									A0907	No Device Output		
4090	2016	TBD	35223	Endocardial/interventricular septal pacing lead	CT1684	Cardiac pacemakers and associated devices			A26	Insufficient Device Problem Information		
4091	2015				CT1684	Cardiac pacemakers and associated devices			A07	Electrical /Electronic Property Problem		
									A0708	Power Problem		
4092	2016	All other reportable incidents	47270	Cardiac resynchronization therapy implantable defibrillator	CT2866	Implantable cardiac electrical stimulators			A09	Output Problem		
									A0907	No Device Output		
4093	2016	TBD			CT257	Implantable cardiac pacemakers			A07	Electrical /Electronic Property Problem		
									A0709	Device Sensing Problem		
4094	2016	TBD			CT257	Implantable cardiac pacemakers			A26	Insufficient Device Problem Information		
4095	2016	All other reportable incidents	47270	Cardiac resynchronization therapy implantable defibrillator	CT2866	Implantable cardiac electrical stimulators			A07	Electrical /Electronic Property Problem		
									A0708	Power Problem		

4096	2015		47270	Cardiac resynchronization therapy implantable defibrillator	CT2866	Implantable cardiac electrical stimulators			A26	Insufficient Device Problem Information		
4097	2015		47270	Cardiac resynchronization therapy implantable defibrillator	CT2866	Implantable cardiac electrical stimulators			A26	Insufficient Device Problem Information		
4098	2015		47270	Cardiac resynchronization therapy implantable defibrillator	CT2866	Implantable cardiac electrical stimulators			A26	Insufficient Device Problem Information		
4099	2016	All other reportable incidents	47270	Cardiac resynchronization therapy implantable defibrillator	CT2866	Implantable cardiac electrical stimulators			A07	Electrical /Electronic Property Problem		
									A0708	Power Problem		
4100	2015				CT257	Implantable cardiac pacemakers			A26	Insufficient Device Problem Information		
4101	2015		35852	Single-chamber implantable defibrillator	CT2866	Implantable cardiac electrical stimulators			A12	Connection Problem	25/09/2015	
									A1204	Failure to Disconnect	25/09/2015	
4102	2016	TBD	47270	Cardiac resynchronization therapy implantable defibrillator	CT2866	Implantable cardiac electrical stimulators			A07	Electrical /Electronic Property Problem		
									A0708	Power Problem		
4103	2015				CT257	Implantable cardiac pacemakers			A07	Electrical /Electronic Property Problem		
									A0709	Device Sensing Problem		
4104	2016	TBD	35223	Endocardial/interventricular septal pacing lead	CT1684	Cardiac pacemakers and associated devices			A15	Activation, Positioning or Separation Problem		
									A150203	Difficult or Delayed Positioning		
4105	2016	TBD	35223	Endocardial/interventricular septal pacing lead	CT1684	Cardiac pacemakers and associated devices			A26	Insufficient Device Problem Information		
4106	2016	TBD			CT257	Implantable cardiac pacemakers			A26	Insufficient Device Problem Information		
4107	2016	All other reportable incidents			CT257	Implantable cardiac pacemakers			A11	Computer Software Problem	09/10/2015	
									A1102	Application Program Problem	09/10/2015	
4108	2016	All other reportable incidents	47270	Cardiac resynchronization therapy implantable defibrillator	CT2866	Implantable cardiac electrical stimulators			A07	Electrical /Electronic Property Problem		
									A0709	Device Sensing Problem		
4109	2016	TBD	47270	Cardiac resynchronization therapy implantable defibrillator	CT2866	Implantable cardiac electrical stimulators			A26	Insufficient Device Problem Information		
4110	2016	All other reportable incidents	47270	Cardiac resynchronization therapy implantable defibrillator	CT2866	Implantable cardiac electrical stimulators			A07	Electrical /Electronic Property Problem	19/11/2016	
									A0708	Power Problem	19/11/2016	
4111	2016	All other reportable incidents	47270	Cardiac resynchronization therapy implantable defibrillator	CT2866	Implantable cardiac electrical stimulators			A07	Electrical /Electronic Property Problem		
									A0708	Power Problem		
4112	2016	TBD			CT257	Implantable cardiac pacemakers			A07	Electrical /Electronic Property Problem		
									A0708	Power Problem		
4113	2016	TBD	47270	Cardiac resynchronization therapy implantable defibrillator	CT2866	Implantable cardiac electrical stimulators			A07	Electrical /Electronic Property Problem	17/02/2016	
									A0708	Power Problem	17/02/2016	
4114	2016	TBD			CT257	Implantable cardiac pacemakers			A07	Electrical /Electronic Property Problem		
									A0708	Power Problem		
4115	2016	TBD			CT257	Implantable cardiac pacemakers			A07	Electrical /Electronic Property Problem		
									A0708	Power Problem		
4116	2016	TBD	47270	Cardiac resynchronization therapy implantable defibrillator	CT2866	Implantable cardiac electrical stimulators			A07	Electrical /Electronic Property Problem		
									A0708	Power Problem		
4117	2015	All other reportable incidents			CT257	Implantable cardiac pacemakers			A16	Protective Measures Problem	01/09/2015	
									A1602	Fail-Safe Problem	01/09/2015	
4118	2016	TBD	47270	Cardiac resynchronization therapy implantable defibrillator	CT2866	Implantable cardiac electrical stimulators			A26	Insufficient Device Problem Information		
4119	2015				CT257	Implantable cardiac pacemakers			A26	Insufficient Device Problem Information		
4120	2016	All other reportable incidents	47270	Cardiac resynchronization therapy implantable defibrillator	CT2866	Implantable cardiac electrical stimulators			A07	Electrical /Electronic Property Problem		
									A0708	Power Problem		
4121	2016	All other reportable incidents	47270	Cardiac resynchronization therapy implantable defibrillator	CT2866	Implantable cardiac electrical stimulators			A07	Electrical /Electronic Property Problem	19/11/2016	
									A0708	Power Problem	19/11/2016	
4122	2016	TBD			CT257	Implantable cardiac pacemakers			A07	Electrical /Electronic Property Problem		
									A0709	Device Sensing Problem		
4123	2016	TBD			CT1684	Cardiac pacemakers and associated devices			A07	Electrical /Electronic Property Problem		
									A0708	Power Problem		
									A09	Output Problem		
									A0907	No Device Output		
4124	2016	TBD	47270	Cardiac resynchronization therapy implantable defibrillator	CT2866	Implantable cardiac electrical stimulators			A07	Electrical /Electronic Property Problem		
									A0708	Power Problem		
4125	2016	All other reportable incidents			CT257	Implantable cardiac pacemakers			A07	Electrical /Electronic Property Problem	05/11/2016	
									A0709	Device Sensing Problem	05/11/2016	
4126	2016	TBD	35223	Endocardial/interventricular septal pacing lead	CT1684	Cardiac pacemakers and associated devices			A07	Electrical /Electronic Property Problem		
									A0709	Device Sensing Problem		
4127	2016	All other reportable incidents	47270	Cardiac resynchronization therapy implantable defibrillator	CT2866	Implantable cardiac electrical stimulators			A07	Electrical /Electronic Property Problem	17/10/2016	
									A0708	Power Problem	17/10/2016	
4128	2016	All other reportable incidents	47270	Cardiac resynchronization therapy implantable defibrillator	CT2866	Implantable cardiac electrical stimulators			A07	Electrical /Electronic Property Problem	19/11/2016	
									A0708	Power Problem	19/11/2016	
4129	2016	All other reportable incidents	47263	Cardiac resynchronization therapy implantable pacemaker	CT1684	Cardiac pacemakers and associated devices			A07	Electrical /Electronic Property Problem	25/01/2016	
									A0708	Power Problem	25/01/2016	
					CT1685	Cardiac pacemakers			A07	Electrical /Electronic Property Problem	25/01/2016	
									A0708	Power Problem	25/01/2016	
					CT257	Implantable cardiac pacemakers			A07	Electrical /Electronic Property Problem	25/01/2016	
									A0708	Power Problem	25/01/2016	
					CT2866	Implantable cardiac electrical stimulators			A07	Electrical /Electronic Property Problem	25/01/2016	
									A0708	Power Problem	25/01/2016	
4130	2016	TBD	47270	Cardiac resynchronization therapy implantable defibrillator	CT2866	Implantable cardiac electrical stimulators			A07	Electrical /Electronic Property Problem		
									A0708	Power Problem		
4131	2016	TBD			CT257	Implantable cardiac pacemakers			A16	Protective Measures Problem		
									A1602	Fail-Safe Problem		
4132	2015	All other reportable incidents	35223	Endocardial/interventricular septal pacing lead	CT1684	Cardiac pacemakers and associated devices			A12	Connection Problem		
4133	2016	All other reportable incidents			CT1684	Cardiac pacemakers and associated devices			A09	Output Problem		
									A0907	No Device Output		
4134	2016	All other reportable incidents			CT1684	Cardiac pacemakers and associated devices			A09	Output Problem		
									A0907	No Device Output		
4135	2015		47270	Cardiac resynchronization therapy implantable defibrillator	CT2866	Implantable cardiac electrical stimulators			A26	Insufficient Device Problem Information		
4136	2016	All other reportable incidents			CT1684	Cardiac pacemakers and associated devices			A09	Output Problem		
									A0907	No Device Output		
4137	2016	TBD			CT257	Implantable cardiac pacemakers			A26	Insufficient Device Problem Information		
4138	2016	All other reportable incidents			CT257	Implantable cardiac pacemakers			A07	Electrical /Electronic Property Problem	19/02/2016	
									A0708	Power Problem	19/02/2016	

4139	2016	All other reportable incidents			CT257	Implantable cardiac pacemakers			A16	Protective Measures Problem	16/05/2016	
									A1602	Fail-Safe Problem	16/05/2016	
4140	2016	All other reportable incidents			CT1684	Cardiac pacemakers and associated devices			A07	Electrical /Electronic Property Problem		
									A0709	Device Sensing Problem		
4141	2016	TBD	47270	Cardiac resynchronization therapy implantable defibrillator	CT2866	Implantable cardiac electrical stimulators			A01	Patient Device Interaction Problem		
									A17	Compatibility Problem		
4142	2016	TBD	47270	Cardiac resynchronization therapy implantable defibrillator	CT2866	Implantable cardiac electrical stimulators			A01	Patient Device Interaction Problem		
									A17	Compatibility Problem		
4143	2016	All other reportable incidents	47270	Cardiac resynchronization therapy implantable defibrillator	CT2866	Implantable cardiac electrical stimulators			A07	Electrical /Electronic Property Problem		
									A0708	Power Problem		
4144	2015				CT257	Implantable cardiac pacemakers			A07	Electrical /Electronic Property Problem		
									A0708	Power Problem		
4145	2016	All other reportable incidents	47270	Cardiac resynchronization therapy implantable defibrillator	CT2866	Implantable cardiac electrical stimulators			A07	Electrical /Electronic Property Problem	18/11/2016	
									A0708	Power Problem	18/11/2016	
4146	2016	TBD	35223	Endocardial/interventricular septal pacing lead	CT1684	Cardiac pacemakers and associated devices			A07	Electrical /Electronic Property Problem		
									A0709	Device Sensing Problem		
4147	2016	TBD	35223	Endocardial/interventricular septal pacing lead	CT1684	Cardiac pacemakers and associated devices			A26	Insufficient Device Problem Information		
4148	2016	All other reportable incidents	35223	Endocardial/interventricular septal pacing lead	CT1684	Cardiac pacemakers and associated devices			A0104	Migration or Expulsion of Device	17/04/2015	
4149	2016	TBD	35223	Endocardial/interventricular septal pacing lead	CT1684	Cardiac pacemakers and associated devices			A26	Insufficient Device Problem Information		
4150	2016	TBD	47270	Cardiac resynchronization therapy implantable defibrillator	CT2866	Implantable cardiac electrical stimulators			A07	Electrical /Electronic Property Problem		
									A0708	Power Problem		
4151	2016	All other reportable incidents	47270	Cardiac resynchronization therapy implantable defibrillator	CT2866	Implantable cardiac electrical stimulators			A07	Electrical /Electronic Property Problem	12/11/2016	
									A0708	Power Problem	12/11/2016	
4152	2016	All other reportable incidents	47270	Cardiac resynchronization therapy implantable defibrillator	CT2866	Implantable cardiac electrical stimulators			A07	Electrical /Electronic Property Problem	14/11/2016	
									A0708	Power Problem	14/11/2016	
4153	2015				CT257	Implantable cardiac pacemakers			A26	Insufficient Device Problem Information		
4154	2015		47270	Cardiac resynchronization therapy implantable defibrillator	CT2866	Implantable cardiac electrical stimulators			A07	Electrical /Electronic Property Problem		
									A0709	Device Sensing Problem		
4155	2016	All other reportable incidents	47270	Cardiac resynchronization therapy implantable defibrillator	CT2866	Implantable cardiac electrical stimulators			A07	Electrical /Electronic Property Problem	13/12/2016	
									A0708	Power Problem	13/12/2016	
4156	2016	TBD	35223	Endocardial/interventricular septal pacing lead	CT1684	Cardiac pacemakers and associated devices			A26	Insufficient Device Problem Information		
4157	2016	All other reportable incidents	47263	Cardiac resynchronization therapy implantable pacemaker	CT1684	Cardiac pacemakers and associated devices			A07	Electrical /Electronic Property Problem		
									A0709	Device Sensing Problem		
					CT1685	Cardiac pacemakers			A07	Electrical /Electronic Property Problem		
									A0709	Device Sensing Problem		
					CT257	Implantable cardiac pacemakers			A07	Electrical /Electronic Property Problem		
									A0709	Device Sensing Problem		
					CT2866	Implantable cardiac electrical stimulators			A07	Electrical /Electronic Property Problem		
									A0709	Device Sensing Problem		
4158	2016	TBD	35223	Endocardial/interventricular septal pacing lead	CT1684	Cardiac pacemakers and associated devices			A26	Insufficient Device Problem Information		
4159	2015		47270	Cardiac resynchronization therapy implantable defibrillator	CT2866	Implantable cardiac electrical stimulators			A07	Electrical /Electronic Property Problem		
									A0708	Power Problem		
4160	2015		47270	Cardiac resynchronization therapy implantable defibrillator	CT2866	Implantable cardiac electrical stimulators			A26	Insufficient Device Problem Information		
4161	2015		47270	Cardiac resynchronization therapy implantable defibrillator	CT2866	Implantable cardiac electrical stimulators			A26	Insufficient Device Problem Information		
4162	2016	All other reportable incidents	47270	Cardiac resynchronization therapy implantable defibrillator	CT2866	Implantable cardiac electrical stimulators			A24	No Apparent Device Problem	01/11/2016	
									A26	Insufficient Device Problem Information	01/11/2016	
4163	2016	All other reportable incidents	47263	Cardiac resynchronization therapy implantable pacemaker	CT1684	Cardiac pacemakers and associated devices			A07	Electrical /Electronic Property Problem	31/12/2015	
									A0708	Power Problem	31/12/2015	
					CT1685	Cardiac pacemakers			A07	Electrical /Electronic Property Problem	31/12/2015	
									A0708	Power Problem	31/12/2015	
					CT257	Implantable cardiac pacemakers			A07	Electrical /Electronic Property Problem	31/12/2015	
									A0708	Power Problem	31/12/2015	
					CT2866	Implantable cardiac electrical stimulators			A07	Electrical /Electronic Property Problem	31/12/2015	
									A0708	Power Problem	31/12/2015	
4164	2016	Death	47270	Cardiac resynchronization therapy implantable defibrillator	CT2866	Implantable cardiac electrical stimulators			A07	Electrical /Electronic Property Problem		
									A0709	Device Sensing Problem		
4165	2016	Death			CT1684	Cardiac pacemakers and associated devices			A09	Output Problem		
									A0907	No Device Output		
4166	2016	TBD	47270	Cardiac resynchronization therapy implantable defibrillator	CT2866	Implantable cardiac electrical stimulators			A07	Electrical /Electronic Property Problem		
									A0708	Power Problem		
4167	2015		35852	Single-chamber implantable defibrillator	CT2866	Implantable cardiac electrical stimulators			A07	Electrical /Electronic Property Problem	27/09/2015	
									A0708	Power Problem	27/09/2015	
4168	2015				CT257	Implantable cardiac pacemakers			A26	Insufficient Device Problem Information		
4169	2015				CT257	Implantable cardiac pacemakers			A07	Electrical /Electronic Property Problem	01/10/2015	
									A0708	Power Problem	01/10/2015	
4170	2016	TBD	35223	Endocardial/interventricular septal pacing lead	CT1684	Cardiac pacemakers and associated devices			A07	Electrical /Electronic Property Problem		
									A0709	Device Sensing Problem		
4171	2016	TBD	47270	Cardiac resynchronization therapy implantable defibrillator	CT2866	Implantable cardiac electrical stimulators			A07	Electrical /Electronic Property Problem		
									A0709	Device Sensing Problem		
4172	2015		47270	Cardiac resynchronization therapy implantable defibrillator	CT2866	Implantable cardiac electrical stimulators			A07	Electrical /Electronic Property Problem		
									A0708	Power Problem		
4173	2016	TBD	35223	Endocardial/interventricular septal pacing lead	CT1684	Cardiac pacemakers and associated devices			A26	Insufficient Device Problem Information		
4174	2016	Unanticipated serious deterioration in state of health	35852	Single-chamber implantable defibrillator	CT2866	Implantable cardiac electrical stimulators			A09	Output Problem		
									A0908	Incorrect, Inadequate or Imprecise Result or Readings		
4175	2016	All other reportable incidents	47270	Cardiac resynchronization therapy implantable defibrillator	CT2866	Implantable cardiac electrical stimulators			A07	Electrical /Electronic Property Problem	21/10/2016	
									A0708	Power Problem	21/10/2016	
4176	2016	All other reportable incidents	47270	Cardiac resynchronization therapy implantable defibrillator	CT2866	Implantable cardiac electrical stimulators			A07	Electrical /Electronic Property Problem		
									A0708	Power Problem		
4177	2016	All other reportable incidents	47270	Cardiac resynchronization therapy implantable defibrillator	CT2866	Implantable cardiac electrical stimulators			A16	Protective Measures Problem		
									A1601	Device Alarm System		
4178	2016	All other reportable incidents	47270	Cardiac resynchronization therapy implantable	CT2866	Implantable cardiac electrical stimulators			A07	Electrical /Electronic Property Problem	19/11/2016	

				defibrillator					A0708	Power Problem	19/11/2016	
4179	2016	TBD	47270	Cardiac resynchronization therapy implantable defibrillator	CT2866	Implantable cardiac electrical stimulators			A07	Electrical /Electronic Property Problem		
									A0709	Device Sensing Problem		
4180	2016	All other reportable incidents	47270	Cardiac resynchronization therapy implantable defibrillator	CT2866	Implantable cardiac electrical stimulators			A07	Electrical /Electronic Property Problem	29/10/2016	
									A0708	Power Problem	29/10/2016	
4181	2016	All other reportable incidents	47270	Cardiac resynchronization therapy implantable defibrillator	CT2866	Implantable cardiac electrical stimulators			A07	Electrical /Electronic Property Problem	01/11/2016	
									A0708	Power Problem	01/11/2016	
4182	2015		47270	Cardiac resynchronization therapy implantable defibrillator	CT2866	Implantable cardiac electrical stimulators			A16	Protective Measures Problem		
									A1602	Fail-Safe Problem		
4183	2016	All other reportable incidents	47270	Cardiac resynchronization therapy implantable defibrillator	CT2866	Implantable cardiac electrical stimulators			A07	Electrical /Electronic Property Problem		
									A0708	Power Problem		
4184	2016	All other reportable incidents	47263	Cardiac resynchronization therapy implantable pacemaker	CT1684	Cardiac pacemakers and associated devices			A07	Electrical /Electronic Property Problem	23/10/2016	
									A0709	Device Sensing Problem	23/10/2016	
					CT1685	Cardiac pacemakers			A07	Electrical /Electronic Property Problem	23/10/2016	
									A0709	Device Sensing Problem	23/10/2016	
					CT257	Implantable cardiac pacemakers			A07	Electrical /Electronic Property Problem	23/10/2016	
									A0709	Device Sensing Problem	23/10/2016	
					CT2866	Implantable cardiac electrical stimulators			A07	Electrical /Electronic Property Problem	23/10/2016	
									A0709	Device Sensing Problem	23/10/2016	
4185	2015	All other reportable incidents	47263	Cardiac resynchronization therapy implantable pacemaker	CT1684	Cardiac pacemakers and associated devices			A16	Protective Measures Problem		
									A1602	Fail-Safe Problem		
					CT1685	Cardiac pacemakers			A16	Protective Measures Problem		
									A1602	Fail-Safe Problem		
					CT257	Implantable cardiac pacemakers			A16	Protective Measures Problem		
									A1602	Fail-Safe Problem		
					CT2866	Implantable cardiac electrical stimulators			A16	Protective Measures Problem		
									A1602	Fail-Safe Problem		
4186	2016	All other reportable incidents	47270	Cardiac resynchronization therapy implantable defibrillator	CT2866	Implantable cardiac electrical stimulators			A07	Electrical /Electronic Property Problem	26/11/2015	
									A0708	Power Problem	26/11/2015	
4187	2016	All other reportable incidents	47263	Cardiac resynchronization therapy implantable pacemaker	CT1684	Cardiac pacemakers and associated devices			A07	Electrical /Electronic Property Problem		
									A0708	Power Problem		
					CT1685	Cardiac pacemakers			A07	Electrical /Electronic Property Problem		
									A0708	Power Problem		
					CT257	Implantable cardiac pacemakers			A07	Electrical /Electronic Property Problem		
									A0708	Power Problem		
					CT2866	Implantable cardiac electrical stimulators			A07	Electrical /Electronic Property Problem		
									A0708	Power Problem		
4188	2016	TBD			CT257	Implantable cardiac pacemakers			A16	Protective Measures Problem		
									A1602	Fail-Safe Problem		
4189	2015				CT257	Implantable cardiac pacemakers			A26	Insufficient Device Problem Information		
4190	2016	All other reportable incidents	47270	Cardiac resynchronization therapy implantable defibrillator	CT2866	Implantable cardiac electrical stimulators			A07	Electrical /Electronic Property Problem	29/10/2016	
									A0708	Power Problem	29/10/2016	
4191	2016	All other reportable incidents	47270	Cardiac resynchronization therapy implantable defibrillator	CT2866	Implantable cardiac electrical stimulators			A24	No Apparent Device Problem	29/11/2016	
									A26	Insufficient Device Problem Information	29/11/2016	
4192	2016	All other reportable incidents	47270	Cardiac resynchronization therapy implantable defibrillator	CT2866	Implantable cardiac electrical stimulators			A07	Electrical /Electronic Property Problem		
									A0708	Power Problem		
4193	2016	All other reportable incidents	47270	Cardiac resynchronization therapy implantable defibrillator	CT2866	Implantable cardiac electrical stimulators			A07	Electrical /Electronic Property Problem	29/11/2016	
									A0708	Power Problem	29/11/2016	
4194	2016	TBD			CT257	Implantable cardiac pacemakers			A26	Insufficient Device Problem Information		
4195	2015		47270	Cardiac resynchronization therapy implantable defibrillator	CT2866	Implantable cardiac electrical stimulators			A26	Insufficient Device Problem Information		
4196	2016	All other reportable incidents	47270	Cardiac resynchronization therapy implantable defibrillator	CT2866	Implantable cardiac electrical stimulators			A07	Electrical /Electronic Property Problem	29/10/2016	
									A0708	Power Problem	29/10/2016	
4197	2016	All other reportable incidents	47270	Cardiac resynchronization therapy implantable defibrillator	CT2866	Implantable cardiac electrical stimulators			A16	Protective Measures Problem		
									A1601	Device Alarm System		
4198	2015		47270	Cardiac resynchronization therapy implantable defibrillator	CT2866	Implantable cardiac electrical stimulators			A07	Electrical /Electronic Property Problem		
									A0708	Power Problem		
4199	2016	All other reportable incidents	47270	Cardiac resynchronization therapy implantable defibrillator	CT2866	Implantable cardiac electrical stimulators			A07	Electrical /Electronic Property Problem	05/12/2016	
									A0708	Power Problem	05/12/2016	
4200	2016	All other reportable incidents	47270	Cardiac resynchronization therapy implantable defibrillator	CT2866	Implantable cardiac electrical stimulators			A07	Electrical /Electronic Property Problem	31/10/2016	
									A0708	Power Problem	31/10/2016	
4201	2016	All other reportable incidents	47270	Cardiac resynchronization therapy implantable defibrillator	CT2866	Implantable cardiac electrical stimulators			A07	Electrical /Electronic Property Problem	30/11/2016	
									A0708	Power Problem	30/11/2016	
4202	2016	All other reportable incidents	47270	Cardiac resynchronization therapy implantable defibrillator	CT2866	Implantable cardiac electrical stimulators			A07	Electrical /Electronic Property Problem	30/11/2016	
									A0708	Power Problem	30/11/2016	
									A26	Insufficient Device Problem Information	30/11/2016	
4203	2016	All other reportable incidents			CT1684	Cardiac pacemakers and associated devices			A09	Output Problem		
									A0907	No Device Output		
4204	2016	All other reportable incidents	47270	Cardiac resynchronization therapy implantable defibrillator	CT2866	Implantable cardiac electrical stimulators			A07	Electrical /Electronic Property Problem		
									A0708	Power Problem		
4205	2016	All other reportable incidents	35852	Single-chamber implantable defibrillator	CT2866	Implantable cardiac electrical stimulators			A07	Electrical /Electronic Property Problem		
									A0709	Device Sensing Problem		
4206	2016	TBD	35223	Endocardial/interventricular septal pacing lead	CT1684	Cardiac pacemakers and associated devices			A0104	Migration or Expulsion of Device		
4207	2015	All other reportable incidents			CT257	Implantable cardiac pacemakers			A07	Electrical /Electronic Property Problem	29/04/2015	
									A0708	Power Problem	29/04/2015	
4208	2016	All other reportable incidents	35223	Endocardial/interventricular septal pacing lead	CT1684	Cardiac pacemakers and associated devices			A26	Insufficient Device Problem Information		
4209	2015				CT257	Implantable cardiac pacemakers			A26	Insufficient Device Problem Information		
4210	2016	All other reportable incidents			CT257	Implantable cardiac pacemakers			A13	Communication or Transmission Problem	04/02/2016	
									A26	Insufficient Device Problem Information	04/02/2016	
4211	2015	All other reportable incidents			CT1684	Cardiac pacemakers and associated devices			A09	Output Problem		
									A0908	Incorrect, Inadequate or Imprecise Result or Readings		
4212	2016	All other reportable incidents	47270	Cardiac resynchronization therapy implantable defibrillator	CT2866	Implantable cardiac electrical stimulators			A07	Electrical /Electronic Property Problem	06/12/2016	
									A0708	Power Problem	06/12/2016	
4213	2016	All other reportable incidents	47263	Cardiac resynchronization therapy implantable pacemaker	CT1684	Cardiac pacemakers and associated devices			A07	Electrical /Electronic Property Problem	16/12/2015	
									A0708	Power Problem	16/12/2015	
					CT1685	Cardiac pacemakers			A07	Electrical /Electronic Property Problem	16/12/2015	

									A0708	Power Problem	16/12/2015	
					CT257	Implantable cardiac pacemakers			A07	Electrical /Electronic Property Problem	16/12/2015	
									A0708	Power Problem	16/12/2015	
					CT2866	Implantable cardiac electrical stimulators			A07	Electrical /Electronic Property Problem	16/12/2015	
									A0708	Power Problem	16/12/2015	
4214	2016	TBD	47270	Cardiac resynchronization therapy implantable defibrillator	CT2866	Implantable cardiac electrical stimulators			A07	Electrical /Electronic Property Problem		
									A0708	Power Problem		
4215	2016	TBD			CT257	Implantable cardiac pacemakers			A07	Electrical /Electronic Property Problem		
									A0709	Device Sensing Problem		
4216	2015	All other reportable incidents			CT257	Implantable cardiac pacemakers			A07	Electrical /Electronic Property Problem	27/08/2015	
									A0708	Power Problem	27/08/2015	
4217	2015	All other reportable incidents			CT257	Implantable cardiac pacemakers			A07	Electrical /Electronic Property Problem		
									A0708	Power Problem		
4218	2016	TBD			CT1684	Cardiac pacemakers and associated devices			A09	Output Problem		
									A0907	No Device Output		
4219	2016	All other reportable incidents	47263	Cardiac resynchronization therapy implantable pacemaker	CT1684	Cardiac pacemakers and associated devices			A07	Electrical /Electronic Property Problem		
									A0708	Power Problem		
					CT1685	Cardiac pacemakers			A07	Electrical /Electronic Property Problem		
									A0708	Power Problem		
					CT257	Implantable cardiac pacemakers			A07	Electrical /Electronic Property Problem		
									A0708	Power Problem		
					CT2866	Implantable cardiac electrical stimulators			A07	Electrical /Electronic Property Problem		
									A0708	Power Problem		
4220	2016	TBD	35223	Endocardial/interventricular septal pacing lead	CT1684	Cardiac pacemakers and associated devices			A26	Insufficient Device Problem Information		
4221	2016	All other reportable incidents	47263	Cardiac resynchronization therapy implantable pacemaker	CT1684	Cardiac pacemakers and associated devices			A26	Insufficient Device Problem Information	08/02/2016	
					CT1685	Cardiac pacemakers			A26	Insufficient Device Problem Information	08/02/2016	
					CT257	Implantable cardiac pacemakers			A26	Insufficient Device Problem Information	08/02/2016	
					CT2866	Implantable cardiac electrical stimulators			A26	Insufficient Device Problem Information	08/02/2016	
4222	2016	Unanticipated serious deterioration in state of health	35223	Endocardial/interventricular septal pacing lead	CT1684	Cardiac pacemakers and associated devices			A07	Electrical /Electronic Property Problem	15/12/2016	
									A0709	Device Sensing Problem	15/12/2016	
4223	2016	All other reportable incidents	47270	Cardiac resynchronization therapy implantable defibrillator	CT2866	Implantable cardiac electrical stimulators			A07	Electrical /Electronic Property Problem		
									A0709	Device Sensing Problem		
4224	2015				CT257	Implantable cardiac pacemakers			A26	Insufficient Device Problem Information		
4225	2016	TBD	35223	Endocardial/interventricular septal pacing lead	CT1684	Cardiac pacemakers and associated devices			A26	Insufficient Device Problem Information		
4226	2015		47270	Cardiac resynchronization therapy implantable defibrillator	CT2866	Implantable cardiac electrical stimulators			A26	Insufficient Device Problem Information		
4227	2015		47270	Cardiac resynchronization therapy implantable defibrillator	CT2866	Implantable cardiac electrical stimulators			A26	Insufficient Device Problem Information		
4228	2015	All other reportable incidents			CT257	Implantable cardiac pacemakers			A07	Electrical /Electronic Property Problem		
									A0708	Power Problem		
4229	2016	All other reportable incidents			CT257	Implantable cardiac pacemakers			A13	Communication or Transmission Problem	04/02/2016	
									A26	Insufficient Device Problem Information	04/02/2016	
4230	2016	All other reportable incidents	47270	Cardiac resynchronization therapy implantable defibrillator	CT2866	Implantable cardiac electrical stimulators			A07	Electrical /Electronic Property Problem	12/12/2016	
									A0708	Power Problem	12/12/2016	
4231	2016	All other reportable incidents	47270	Cardiac resynchronization therapy implantable defibrillator	CT2866	Implantable cardiac electrical stimulators			A07	Electrical /Electronic Property Problem	29/10/2016	
									A0708	Power Problem	29/10/2016	
4232	2015				CT257	Implantable cardiac pacemakers			A07	Electrical /Electronic Property Problem		
									A0709	Device Sensing Problem		
4233	2015		47270	Cardiac resynchronization therapy implantable defibrillator	CT2866	Implantable cardiac electrical stimulators			A07	Electrical /Electronic Property Problem		
									A0708	Power Problem		
4234	2016	TBD	47270	Cardiac resynchronization therapy implantable defibrillator	CT2866	Implantable cardiac electrical stimulators			A07	Electrical /Electronic Property Problem		
									A0709	Device Sensing Problem		
4235	2016	All other reportable incidents	47270	Cardiac resynchronization therapy implantable defibrillator	CT2866	Implantable cardiac electrical stimulators			A07	Electrical /Electronic Property Problem		
									A0708	Power Problem		
4236	2016	All other reportable incidents	47270	Cardiac resynchronization therapy implantable defibrillator	CT2866	Implantable cardiac electrical stimulators			A26	Insufficient Device Problem Information		
4237	2016	All other reportable incidents	47270	Cardiac resynchronization therapy implantable defibrillator	CT2866	Implantable cardiac electrical stimulators			A07	Electrical /Electronic Property Problem	29/11/2016	
									A0708	Power Problem	29/11/2016	
4238	2015	All other reportable incidents	47270	Cardiac resynchronization therapy implantable defibrillator	CT2866	Implantable cardiac electrical stimulators			A13	Communication or Transmission Problem	22/09/2015	
									A26	Insufficient Device Problem Information	22/09/2015	
4239	2015	All other reportable incidents	47270	Cardiac resynchronization therapy implantable defibrillator	CT2866	Implantable cardiac electrical stimulators			A07	Electrical /Electronic Property Problem	18/09/2015	
									A0708	Power Problem	18/09/2015	
4240	2016	TBD	47270	Cardiac resynchronization therapy implantable defibrillator	CT2866	Implantable cardiac electrical stimulators			A07	Electrical /Electronic Property Problem		
									A0708	Power Problem		
4241	2016	TBD	35223	Endocardial/interventricular septal pacing lead	CT1684	Cardiac pacemakers and associated devices			A07	Electrical /Electronic Property Problem		
									A0709	Device Sensing Problem		
4242	2015		47270	Cardiac resynchronization therapy implantable defibrillator	CT2866	Implantable cardiac electrical stimulators			A07	Electrical /Electronic Property Problem		
									A0708	Power Problem		
4243	2016	All other reportable incidents	47270	Cardiac resynchronization therapy implantable defibrillator	CT2866	Implantable cardiac electrical stimulators			A07	Electrical /Electronic Property Problem		
									A0708	Power Problem		
4244	2016	TBD			CT257	Implantable cardiac pacemakers			A26	Insufficient Device Problem Information		
4245	2015		47270	Cardiac resynchronization therapy implantable defibrillator	CT2866	Implantable cardiac electrical stimulators			A26	Insufficient Device Problem Information		
4246	2016	All other reportable incidents			CT257	Implantable cardiac pacemakers			A21	Labeling, Instructions for Use or Training Problem		
4247	2016	All other reportable incidents			CT257	Implantable cardiac pacemakers			A16	Protective Measures Problem		
									A1602	Fail-Safe Problem		
4248	2016	All other reportable incidents			CT257	Implantable cardiac pacemakers			A07	Electrical /Electronic Property Problem	07/03/2016	
									A0708	Power Problem	07/03/2016	
4249	2015	All other reportable incidents			CT257	Implantable cardiac pacemakers			A07	Electrical /Electronic Property Problem	25/09/2015	
									A0708	Power Problem	25/09/2015	
4250	2015	All other reportable incidents	47270	Cardiac resynchronization therapy implantable defibrillator	CT2866	Implantable cardiac electrical stimulators			A16	Protective Measures Problem	09/10/2015	
									A1602	Fail-Safe Problem	09/10/2015	
4251	2015	All other reportable incidents			CT257	Implantable cardiac pacemakers			A07	Electrical /Electronic Property Problem	01/10/2015	
									A0708	Power Problem	01/10/2015	
4252	2016	TBD			CT257	Implantable cardiac pacemakers			A26	Insufficient Device Problem Information		
4253	2016	TBD	35223	Endocardial/interventricular septal pacing lead	CT1684	Cardiac pacemakers and associated devices			A26	Insufficient Device Problem Information		

4254	2015				CT257	Implantable cardiac pacemakers			A26	Insufficient Device Problem Information		
4255	2016	TBD	47270	Cardiac resynchronization therapy implantable defibrillator	CT2866	Implantable cardiac electrical stimulators			A07	Electrical /Electronic Property Problem		
									A0708	Power Problem		
4256	2016	TBD			CT257	Implantable cardiac pacemakers			A26	Insufficient Device Problem Information		
4257	2016	All other reportable incidents	47270	Cardiac resynchronization therapy implantable defibrillator	CT2866	Implantable cardiac electrical stimulators			A07	Electrical /Electronic Property Problem	19/11/2016	
									A0708	Power Problem	19/11/2016	
4258	2016	All other reportable incidents			CT1684	Cardiac pacemakers and associated devices			A07	Electrical /Electronic Property Problem		
									A0708	Power Problem		
4259	2016	All other reportable incidents	47270	Cardiac resynchronization therapy implantable defibrillator	CT2866	Implantable cardiac electrical stimulators			A07	Electrical /Electronic Property Problem		
									A0708	Power Problem		
4260	2016	All other reportable incidents			CT1684	Cardiac pacemakers and associated devices			A09	Output Problem		
									A0908	Incorrect, Inadequate or Imprecise Result or Readings		
4261	2016	All other reportable incidents	47263	Cardiac resynchronization therapy implantable pacemaker	CT1684	Cardiac pacemakers and associated devices			A07	Electrical /Electronic Property Problem	03/12/2015	
									A0708	Power Problem	03/12/2015	
					CT1685	Cardiac pacemakers			A07	Electrical /Electronic Property Problem	03/12/2015	
									A0708	Power Problem	03/12/2015	
					CT257	Implantable cardiac pacemakers			A07	Electrical /Electronic Property Problem	03/12/2015	
									A0708	Power Problem	03/12/2015	
					CT2866	Implantable cardiac electrical stimulators			A07	Electrical /Electronic Property Problem	03/12/2015	
									A0708	Power Problem	03/12/2015	
4262	2015	All other reportable incidents			CT257	Implantable cardiac pacemakers			A07	Electrical /Electronic Property Problem		
									A0708	Power Problem		
4263	2015				CT257	Implantable cardiac pacemakers			A09	Output Problem		
									A0907	No Device Output		
4264	2015		47270	Cardiac resynchronization therapy implantable defibrillator	CT2866	Implantable cardiac electrical stimulators			A26	Insufficient Device Problem Information		
4265	2015				CT257	Implantable cardiac pacemakers			A26	Insufficient Device Problem Information		
4266	2015		47270	Cardiac resynchronization therapy implantable defibrillator	CT2866	Implantable cardiac electrical stimulators			A26	Insufficient Device Problem Information		
4267	2015				CT1684	Cardiac pacemakers and associated devices			A10	Temperature Problem		
									A1005	Overheating of Device		
									A1009	Smoking		
4268	2015		47270	Cardiac resynchronization therapy implantable defibrillator	CT2866	Implantable cardiac electrical stimulators			A07	Electrical /Electronic Property Problem		
									A0708	Power Problem		
4269	2015				CT1684	Cardiac pacemakers and associated devices			A07	Electrical /Electronic Property Problem		
									A0708	Power Problem		
4270	2016	Death	47270	Cardiac resynchronization therapy implantable defibrillator	CT2866	Implantable cardiac electrical stimulators			A26	Insufficient Device Problem Information	26/02/2016	
4271	2016	All other reportable incidents			CT257	Implantable cardiac pacemakers			A07	Electrical /Electronic Property Problem	15/09/2015	
									A0708	Power Problem	15/09/2015	
4272	2015	All other reportable incidents	47270	Cardiac resynchronization therapy implantable defibrillator	CT2866	Implantable cardiac electrical stimulators			A07	Electrical /Electronic Property Problem		
									A0708	Power Problem		
4273	2016	TBD	35223	Endocardial/interventricular septal pacing lead	CT1684	Cardiac pacemakers and associated devices			A26	Insufficient Device Problem Information		
4274	2016	All other reportable incidents	47270	Cardiac resynchronization therapy implantable defibrillator	CT2866	Implantable cardiac electrical stimulators			A1104	Computer Operating System Problem		
4275	2016	TBD			CT257	Implantable cardiac pacemakers			A09	Output Problem		
									A0908	Incorrect, Inadequate or Imprecise Result or Readings		
4276	2016	TBD	35223	Endocardial/interventricular septal pacing lead	CT1684	Cardiac pacemakers and associated devices			A26	Insufficient Device Problem Information		
4277	2016	All other reportable incidents	47270	Cardiac resynchronization therapy implantable defibrillator	CT2866	Implantable cardiac electrical stimulators			A07	Electrical /Electronic Property Problem		
									A0708	Power Problem		
4278	2016	TBD	47270	Cardiac resynchronization therapy implantable defibrillator	CT2866	Implantable cardiac electrical stimulators			A07	Electrical /Electronic Property Problem		
									A0708	Power Problem		
4279	2016	Unanticipated serious deterioration in state of health	35223	Endocardial/interventricular septal pacing lead	CT1684	Cardiac pacemakers and associated devices			A15	Activation, Positioning or Separation Problem		
									A150203	Difficult or Delayed Positioning		
4280	2015		47270	Cardiac resynchronization therapy implantable defibrillator	CT2866	Implantable cardiac electrical stimulators			A15	Activation, Positioning or Separation Problem	05/10/2015	
									A150101	Activation Failure	05/10/2015	
4281	2015				CT257	Implantable cardiac pacemakers			A11	Computer Software Problem		
									A1102	Application Program Problem		
4282	2016	TBD			CT257	Implantable cardiac pacemakers			A07	Electrical /Electronic Property Problem		
									A0708	Power Problem		
									A0709	Device Sensing Problem		
4283	2015	All other reportable incidents			CT257	Implantable cardiac pacemakers			A1104	Computer Operating System Problem		
4284	2016	All other reportable incidents	47263	Cardiac resynchronization therapy implantable pacemaker	CT1684	Cardiac pacemakers and associated devices			A07	Electrical /Electronic Property Problem		
									A0708	Power Problem		
					CT1685	Cardiac pacemakers			A07	Electrical /Electronic Property Problem		
									A0708	Power Problem		
					CT257	Implantable cardiac pacemakers			A07	Electrical /Electronic Property Problem		
									A0708	Power Problem		
					CT2866	Implantable cardiac electrical stimulators			A07	Electrical /Electronic Property Problem		
									A0708	Power Problem		
4285	2016	All other reportable incidents			CT1684	Cardiac pacemakers and associated devices			A11	Computer Software Problem		
									A1102	Application Program Problem		
4286	2016	All other reportable incidents	35223	Endocardial/interventricular septal pacing lead	CT1684	Cardiac pacemakers and associated devices			A07	Electrical /Electronic Property Problem		
									A0709	Device Sensing Problem		
4287	2015	All other reportable incidents			CT257	Implantable cardiac pacemakers			A07	Electrical /Electronic Property Problem		
									A0708	Power Problem		
4288	2016	All other reportable incidents			CT1684	Cardiac pacemakers and associated devices			A09	Output Problem		
									A0907	No Device Output		
4289	2016	TBD			CT257	Implantable cardiac pacemakers			A07	Electrical /Electronic Property Problem		
									A0708	Power Problem		
4290	2015		47270	Cardiac resynchronization therapy implantable defibrillator	CT2866	Implantable cardiac electrical stimulators			A26	Insufficient Device Problem Information		
4291	2016	All other reportable incidents	47263	Cardiac resynchronization therapy implantable pacemaker	CT1684	Cardiac pacemakers and associated devices			A09	Output Problem		
									A0907	No Device Output		
					CT1685	Cardiac pacemakers			A09	Output Problem		
									A0907	No Device Output		

					CT257	Implantable cardiac pacemakers			A09	Output Problem		
									A0907	No Device Output		
					CT2866	Implantable cardiac electrical stimulators			A09	Output Problem		
									A0907	No Device Output		
4292	2015		47270	Cardiac resynchronization therapy implantable defibrillator	CT2866	Implantable cardiac electrical stimulators			A26	Insufficient Device Problem Information		
4293	2015				CT257	Implantable cardiac pacemakers			A26	Insufficient Device Problem Information		
4294	2016	TBD	47270	Cardiac resynchronization therapy implantable defibrillator	CT2866	Implantable cardiac electrical stimulators			A07	Electrical /Electronic Property Problem		
									A0708	Power Problem		
4295	2016	Unanticipated serious deterioration in state of health			CT1684	Cardiac pacemakers and associated devices			A07	Electrical /Electronic Property Problem		
									A0709	Device Sensing Problem		
									A09	Output Problem		
									A0908	Incorrect, Inadequate or Imprecise Result or Readings		
4296	2016	All other reportable incidents			CT257	Implantable cardiac pacemakers			A13	Communication or Transmission Problem	30/11/2016	
									A26	Insufficient Device Problem Information	30/11/2016	
4297	2016	TBD	47263	Cardiac resynchronization therapy implantable pacemaker	CT1684	Cardiac pacemakers and associated devices			A07	Electrical /Electronic Property Problem		
									A0709	Device Sensing Problem		
					CT1685	Cardiac pacemakers			A07	Electrical /Electronic Property Problem		
									A0709	Device Sensing Problem		
					CT257	Implantable cardiac pacemakers			A07	Electrical /Electronic Property Problem		
									A0709	Device Sensing Problem		
					CT2866	Implantable cardiac electrical stimulators			A07	Electrical /Electronic Property Problem		
									A0709	Device Sensing Problem		
4298	2016	TBD	47270	Cardiac resynchronization therapy implantable defibrillator	CT2866	Implantable cardiac electrical stimulators			A07	Electrical /Electronic Property Problem		
									A0708	Power Problem		
4299	2016	All other reportable incidents	35223	Endocardial/interventricular septal pacing lead	CT1684	Cardiac pacemakers and associated devices			A26	Insufficient Device Problem Information		
4300	2016	TBD	47270	Cardiac resynchronization therapy implantable defibrillator	CT2866	Implantable cardiac electrical stimulators			A17	Compatibility Problem		
									A1701	Component or Accessory Incompatibility		
4301	2016	All other reportable incidents			CT257	Implantable cardiac pacemakers			A07	Electrical /Electronic Property Problem		
									A0708	Power Problem		
4302	2018	All other reportable incidents			CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A07	Electrical /Electronic Property Problem		
									A0709	Device Sensing Problem		
4303	2015				CT1684	Cardiac pacemakers and associated devices			A09	Output Problem		
									A0907	No Device Output		
4304	2018	All other reportable incidents			CT257	Implantable cardiac pacemakers	F1001	Absence of Treatment	A07	Electrical /Electronic Property Problem		
									A0708	Power Problem		
							F1905	Device Revision or Replacement	A07	Electrical /Electronic Property Problem		
									A0708	Power Problem		
4305	2018	All other reportable incidents			CT257	Implantable cardiac pacemakers	F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A07	Electrical /Electronic Property Problem		
									A0709	Device Sensing Problem		
4306	2018	All other reportable incidents	35852	Single-chamber implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A07	Electrical /Electronic Property Problem		
									A0709	Device Sensing Problem		
4307	2018	All other reportable incidents	47270	Cardiac resynchronization therapy implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	F1905	Device Revision or Replacement	A07	Electrical /Electronic Property Problem	14/06/2018	
									A0709	Device Sensing Problem	14/06/2018	
4308	2016	All other reportable incidents	47263	Cardiac resynchronization therapy implantable pacemaker	CT1684	Cardiac pacemakers and associated devices			A07	Electrical /Electronic Property Problem		
									A0708	Power Problem		
					CT1685	Cardiac pacemakers			A07	Electrical /Electronic Property Problem		
									A0708	Power Problem		
					CT257	Implantable cardiac pacemakers			A07	Electrical /Electronic Property Problem		
									A0708	Power Problem		
					CT2866	Implantable cardiac electrical stimulators			A07	Electrical /Electronic Property Problem		
									A0708	Power Problem		
4309	2015				CT257	Implantable cardiac pacemakers			A07	Electrical /Electronic Property Problem		
									A0708	Power Problem		
4310	2015				CT257	Implantable cardiac pacemakers			A07	Electrical /Electronic Property Problem		
									A0708	Power Problem		
4311	2015	All other reportable incidents			CT257	Implantable cardiac pacemakers			A07	Electrical /Electronic Property Problem	06/07/2015	
									A0708	Power Problem	06/07/2015	
4312	2018	All other reportable incidents	47270	Cardiac resynchronization therapy implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	F1905	Device Revision or Replacement	A07	Electrical /Electronic Property Problem		
									A0708	Power Problem		
4313	2015		47270	Cardiac resynchronization therapy implantable defibrillator	CT2866	Implantable cardiac electrical stimulators			A07	Electrical /Electronic Property Problem		
									A0708	Power Problem		
4314	2018	All other reportable incidents	47270	Cardiac resynchronization therapy implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A07	Electrical /Electronic Property Problem	06/04/2018	
									A0708	Power Problem	06/04/2018	
4315	2015				CT257	Implantable cardiac pacemakers			A07	Electrical /Electronic Property Problem	01/06/2015	
									A0708	Power Problem	01/06/2015	
4316	2017	All other reportable incidents	47270	Cardiac resynchronization therapy implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A07	Electrical /Electronic Property Problem		
									A0708	Power Problem		
4317	2015				CT257	Implantable cardiac pacemakers			A07	Electrical /Electronic Property Problem		
									A0709	Device Sensing Problem		
4318	2018	All other reportable incidents	47263	Cardiac resynchronization therapy implantable pacemaker	CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A07	Electrical /Electronic Property Problem		
									A0709	Device Sensing Problem		
					CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A07	Electrical /Electronic Property Problem		
									A0709	Device Sensing Problem		
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A07	Electrical /Electronic Property Problem		
									A0709	Device Sensing Problem		
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A07	Electrical /Electronic Property Problem		
									A0709	Device Sensing Problem		
4319	2016	TBD			CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A07	Electrical /Electronic Property Problem		
									A0709	Device Sensing Problem		
4320	2018	All other reportable incidents	47270	Cardiac resynchronization therapy implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	F1905	Device Revision or Replacement	A07	Electrical /Electronic Property Problem		
									A0708	Power Problem		
4321	2018	Unanticipated serious deterioration in state of health	35852	Single-chamber implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E2401	Insufficient Information	A26	Insufficient Device Problem Information		
4322	2017	TBD	47270	Cardiac resynchronization therapy implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A16	Protective Measures Problem		
									A1602	Fail-Safe Problem		
4323	2017	All other reportable incidents	47270	Cardiac resynchronization therapy implantable	CT2866	Implantable cardiac electrical stimulators	F1905	Device Revision or Replacement	A07	Electrical /Electronic Property Problem	08/12/2016	

				defibrillator					A0708	Power Problem	08/12/2016	
4324	2017	All other reportable incidents	47270	Cardiac resynchronization therapy implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	F1905	Device Revision or Replacement	A07	Electrical /Electronic Property Problem	01/12/2016	
									A0708	Power Problem	01/12/2016	
4325	2017	All other reportable incidents	47270	Cardiac resynchronization therapy implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	F1905	Device Revision or Replacement	A07	Electrical /Electronic Property Problem	06/12/2016	
									A0708	Power Problem	06/12/2016	
4326	2015		35852	Single-chamber implantable defibrillator	CT2866	Implantable cardiac electrical stimulators			A07	Electrical /Electronic Property Problem		
									A0708	Power Problem		
4327	2018	All other reportable incidents			CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A07	Electrical /Electronic Property Problem		
									A0709	Device Sensing Problem		
4328	2017	All other reportable incidents	47270	Cardiac resynchronization therapy implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	F1905	Device Revision or Replacement	A07	Electrical /Electronic Property Problem	20/12/2016	
									A0708	Power Problem	20/12/2016	
4329	2017	All other reportable incidents	47270	Cardiac resynchronization therapy implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	F1905	Device Revision or Replacement	A07	Electrical /Electronic Property Problem	17/12/2016	
									A0708	Power Problem	17/12/2016	
4330	2018	All other reportable incidents	47270	Cardiac resynchronization therapy implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A07	Electrical /Electronic Property Problem		
									A0709	Device Sensing Problem		
4331	2017	All other reportable incidents	47270	Cardiac resynchronization therapy implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	F1905	Device Revision or Replacement	A07	Electrical /Electronic Property Problem		
									A0708	Power Problem		
4332	2017	All other reportable incidents	47270	Cardiac resynchronization therapy implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	F1905	Device Revision or Replacement	A07	Electrical /Electronic Property Problem	15/12/2016	
									A0708	Power Problem	15/12/2016	
4333	2017	All other reportable incidents	47270	Cardiac resynchronization therapy implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	F1905	Device Revision or Replacement	A07	Electrical /Electronic Property Problem		
									A0708	Power Problem		
4334	2017	All other reportable incidents	47270	Cardiac resynchronization therapy implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	F1905	Device Revision or Replacement	A07	Electrical /Electronic Property Problem	12/12/2016	
									A0708	Power Problem	12/12/2016	
4335	2018	All other reportable incidents			CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A07	Electrical /Electronic Property Problem		
									A0709	Device Sensing Problem		
4336	2018	All other reportable incidents	47270	Cardiac resynchronization therapy implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	F1905	Device Revision or Replacement	A07	Electrical /Electronic Property Problem	07/07/2018	
									A0708	Power Problem	07/07/2018	
4337	2017	TBD	35223	Endocardial/interventricular septal pacing lead	CT1684	Cardiac pacemakers and associated devices	F1905	Device Revision or Replacement	A07	Electrical /Electronic Property Problem		
									A0709	Device Sensing Problem		
									A26	Insufficient Device Problem Information		
4338	2015				CT257	Implantable cardiac pacemakers			A07	Electrical /Electronic Property Problem		
									A0708	Power Problem		
4339	2018	All other reportable incidents			CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A13	Communication or Transmission Problem		
									A26	Insufficient Device Problem Information		
4340	2018	All other reportable incidents	47263	Cardiac resynchronization therapy implantable pacemaker	CT1684	Cardiac pacemakers and associated devices	E2007	Fall	A07	Electrical /Electronic Property Problem		
									A0709	Device Sensing Problem		
							F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A07	Electrical /Electronic Property Problem		
									A0709	Device Sensing Problem		
					CT1685	Cardiac pacemakers	E2007	Fall	A07	Electrical /Electronic Property Problem		
									A0709	Device Sensing Problem		
							F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A07	Electrical /Electronic Property Problem		
									A0709	Device Sensing Problem		
					CT257	Implantable cardiac pacemakers	E2007	Fall	A07	Electrical /Electronic Property Problem		
									A0709	Device Sensing Problem		
							F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A07	Electrical /Electronic Property Problem		
									A0709	Device Sensing Problem		
					CT2866	Implantable cardiac electrical stimulators	E2007	Fall	A07	Electrical /Electronic Property Problem		
									A0709	Device Sensing Problem		
							F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A07	Electrical /Electronic Property Problem		
									A0709	Device Sensing Problem		
4341	2017	All other reportable incidents	47270	Cardiac resynchronization therapy implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	F1905	Device Revision or Replacement	A07	Electrical /Electronic Property Problem	31/10/2016	
									A0708	Power Problem	31/10/2016	
4342	2017	TBD	47270	Cardiac resynchronization therapy implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A07	Electrical /Electronic Property Problem		
									A0709	Device Sensing Problem		
4343	2017	All other reportable incidents	47270	Cardiac resynchronization therapy implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A07	Electrical /Electronic Property Problem	09/12/2016	
									A0708	Power Problem	09/12/2016	
4344	2017	All other reportable incidents	47270	Cardiac resynchronization therapy implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	F1905	Device Revision or Replacement	A07	Electrical /Electronic Property Problem	10/12/2016	
									A0708	Power Problem	10/12/2016	
4345	2017	All other reportable incidents	47270	Cardiac resynchronization therapy implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	F1905	Device Revision or Replacement	A07	Electrical /Electronic Property Problem	09/12/2016	
									A0708	Power Problem	09/12/2016	
4346	2017	TBD	35852	Single-chamber implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A09	Output Problem		
									A0907	No Device Output		
4347	2017	TBD			CT257	Implantable cardiac pacemakers	F05	Delay to Treatment/ Therapy	A15	Activation, Positioning or Separation Problem		
									A150101	Activation Failure		
4348	2016	All other reportable incidents	47270	Cardiac resynchronization therapy implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	F1905	Device Revision or Replacement	A24	No Apparent Device Problem	23/11/2016	
									A26	Insufficient Device Problem Information	23/11/2016	
4349	2017	All other reportable incidents	47263	Cardiac resynchronization therapy implantable pacemaker	CT1684	Cardiac pacemakers and associated devices	F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A07	Electrical /Electronic Property Problem		
									A0709	Device Sensing Problem		
					CT1685	Cardiac pacemakers	F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A07	Electrical /Electronic Property Problem		
									A0709	Device Sensing Problem		
					CT257	Implantable cardiac pacemakers	F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A07	Electrical /Electronic Property Problem		
									A0709	Device Sensing Problem		
					CT2866	Implantable cardiac electrical stimulators	F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A07	Electrical /Electronic Property Problem		
									A0709	Device Sensing Problem		
4350	2017	All other reportable incidents			CT257	Implantable cardiac pacemakers	F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A1304	Telemetry Discrepancy		
									A26	Insufficient Device Problem Information		
4351	2017	All other reportable incidents	47270	Cardiac resynchronization therapy implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A07	Electrical /Electronic Property Problem	23/10/2016	
									A0709	Device Sensing Problem	23/10/2016	
							F1905	Device Revision or Replacement	A07	Electrical /Electronic Property Problem	23/10/2016	
									A0709	Device Sensing Problem	23/10/2016	
4352	2018	All other reportable incidents			CT257	Implantable cardiac pacemakers	F1905	Device Revision or Replacement	A07	Electrical /Electronic Property Problem	27/07/2018	
									A0708	Power Problem	27/07/2018	
4353	2017	TBD	35223	Endocardial/interventricular septal pacing lead	CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A05	Mechanical Problem		
									A0510	Retraction Problem		
									A07	Electrical /Electronic Property Problem		
									A0709	Device Sensing Problem		
4354	2017	TBD	35223	Endocardial/interventricular septal pacing lead	CT1684	Cardiac pacemakers and associated devices			A26	Insufficient Device Problem Information		

4355	2018	All other reportable incidents	47270	Cardiac resynchronization therapy implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A07	Electrical /Electronic Property Problem		
									A0709	Device Sensing Problem		
4356	2018	All other reportable incidents	47270	Cardiac resynchronization therapy implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A07	Electrical /Electronic Property Problem		
									A0708	Power Problem		
4357	2017	TBD	47270	Cardiac resynchronization therapy implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A07	Electrical /Electronic Property Problem		
									A0708	Power Problem		
4358	2018	All other reportable incidents			CT257	Implantable cardiac pacemakers	F1905	Device Revision or Replacement	A07	Electrical /Electronic Property Problem	14/06/2018	
									A0709	Device Sensing Problem	14/06/2018	
4359	2018	All other reportable incidents	47263	Cardiac resynchronization therapy implantable pacemaker	CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A07	Electrical /Electronic Property Problem		
									A0709	Device Sensing Problem		
					CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A07	Electrical /Electronic Property Problem		
									A0709	Device Sensing Problem		
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A07	Electrical /Electronic Property Problem		
									A0709	Device Sensing Problem		
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A07	Electrical /Electronic Property Problem		
									A0709	Device Sensing Problem		
4360	2017	All other reportable incidents	47270	Cardiac resynchronization therapy implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	F1905	Device Revision or Replacement	A07	Electrical /Electronic Property Problem	13/12/2016	
									A0708	Power Problem	13/12/2016	
4361	2017	All other reportable incidents	47270	Cardiac resynchronization therapy implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	F1905	Device Revision or Replacement	A07	Electrical /Electronic Property Problem	09/12/2016	
									A0708	Power Problem	09/12/2016	
4362	2017	All other reportable incidents	47270	Cardiac resynchronization therapy implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	F1905	Device Revision or Replacement	A07	Electrical /Electronic Property Problem	21/12/2016	
									A0708	Power Problem	21/12/2016	
4363	2017	All other reportable incidents	47270	Cardiac resynchronization therapy implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	F1905	Device Revision or Replacement	A24	No Apparent Device Problem	04/01/2017	
									A26	Insufficient Device Problem Information	04/01/2017	
4364	2018	All other reportable incidents	47270	Cardiac resynchronization therapy implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A07	Electrical /Electronic Property Problem		
									A0709	Device Sensing Problem		
4365	2017	All other reportable incidents	47270	Cardiac resynchronization therapy implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	F1905	Device Revision or Replacement	A07	Electrical /Electronic Property Problem	07/12/2016	
									A0708	Power Problem	07/12/2016	
4366	2017	All other reportable incidents	35852	Single-chamber implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A0104	Migration or Expulsion of Device		
4367	2018	All other reportable incidents			CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A07	Electrical /Electronic Property Problem		
									A0708	Power Problem		
4368	2017	TBD	35223	Endocardial/interventricular septal pacing lead	CT1684	Cardiac pacemakers and associated devices	F1905	Device Revision or Replacement	A05	Mechanical Problem		
									A051201	Device Dislodged or Dislocated		
4369	2018	All other reportable incidents			CT257	Implantable cardiac pacemakers	F1905	Device Revision or Replacement	A07	Electrical /Electronic Property Problem	03/08/2018	
									A0708	Power Problem	03/08/2018	
4370	2017	TBD	35223	Endocardial/interventricular septal pacing lead	CT1684	Cardiac pacemakers and associated devices	E0402	Hypersensitivity/Allergic Reaction	A26	Insufficient Device Problem Information		
4371	2018	All other reportable incidents			CT1684	Cardiac pacemakers and associated devices	F1001	Absence of Treatment	A09	Output Problem		
									A0907	No Device Output		
4372	2018	All other reportable incidents			CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A09	Output Problem		
									A0907	No Device Output		
4373	2018	All other reportable incidents	35223	Endocardial/interventricular septal pacing lead	CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A15	Activation, Positioning or Separation Problem		
									A150203	Difficult or Delayed Positioning		
4374	2018	All other reportable incidents	35852	Single-chamber implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A07	Electrical /Electronic Property Problem		
									A0709	Device Sensing Problem		
4375	2018	All other reportable incidents	35223	Endocardial/interventricular septal pacing lead	CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A0104	Migration or Expulsion of Device		
4376	2018	All other reportable incidents	35852	Single-chamber implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A1104	Computer Operating System Problem		
4377	2018	All other reportable incidents			CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A07	Electrical /Electronic Property Problem		
									A0709	Device Sensing Problem		
4378	2017	All other reportable incidents			CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A1104	Computer Operating System Problem		
4379	2017	All other reportable incidents	47270	Cardiac resynchronization therapy implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	F1905	Device Revision or Replacement	A07	Electrical /Electronic Property Problem	13/12/2016	
									A0708	Power Problem	13/12/2016	
4380	2017	All other reportable incidents	47270	Cardiac resynchronization therapy implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	F1905	Device Revision or Replacement	A07	Electrical /Electronic Property Problem	05/01/2017	
									A0708	Power Problem	05/01/2017	
4381	2017	All other reportable incidents	35223	Endocardial/interventricular septal pacing lead	CT1684	Cardiac pacemakers and associated devices	F1905	Device Revision or Replacement	A07	Electrical /Electronic Property Problem	09/12/2016	
									A0709	Device Sensing Problem	09/12/2016	
4382	2018	All other reportable incidents	35039	Epicardial pacing lead	CT1684	Cardiac pacemakers and associated devices	F1905	Device Revision or Replacement	A07	Electrical /Electronic Property Problem	30/10/2017	
									A0709	Device Sensing Problem	30/10/2017	
4383	2017	All other reportable incidents	35223	Endocardial/interventricular septal pacing lead	CT1684	Cardiac pacemakers and associated devices	F1905	Device Revision or Replacement	A07	Electrical /Electronic Property Problem	09/12/2016	
									A0709	Device Sensing Problem	09/12/2016	
4384	2018	All other reportable incidents	47270	Cardiac resynchronization therapy implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	F1905	Device Revision or Replacement	A04	Material Integrity Problem	02/07/2018	
									A0413	Material Separation	02/07/2018	
4385	2017	All other reportable incidents	47270	Cardiac resynchronization therapy implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	F1905	Device Revision or Replacement	A07	Electrical /Electronic Property Problem		
									A0708	Power Problem		
4386	2018	All other reportable incidents	35223	Endocardial/interventricular septal pacing lead	CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A05	Mechanical Problem	14/08/2018	
									A0510	Retraction Problem	14/08/2018	
4387	2017	All other reportable incidents	47270	Cardiac resynchronization therapy implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	F1905	Device Revision or Replacement	A07	Electrical /Electronic Property Problem	07/12/2016	
									A0708	Power Problem	07/12/2016	
4388	2018	All other reportable incidents	47270	Cardiac resynchronization therapy implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A07	Electrical /Electronic Property Problem		
									A0709	Device Sensing Problem		
4389	2017	TBD	47270	Cardiac resynchronization therapy implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A07	Electrical /Electronic Property Problem		
									A0708	Power Problem		
4390	2017	All other reportable incidents	35852	Single-chamber implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A09	Output Problem		
									A090404	Output Above Specifications		
4391	2017	TBD	35223	Endocardial/interventricular septal pacing lead	CT1684	Cardiac pacemakers and associated devices	F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A07	Electrical /Electronic Property Problem		
									A0709	Device Sensing Problem		
									A09	Output Problem		
									A0907	No Device Output		
4392	2018	All other reportable incidents	47270	Cardiac resynchronization therapy implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A07	Electrical /Electronic Property Problem		
									A0709	Device Sensing Problem		
4393	2018	All other reportable incidents	35223	Endocardial/interventricular septal pacing lead	CT1684	Cardiac pacemakers and associated devices	F14	Prolonged Episode of Care	A07	Electrical /Electronic Property Problem		
									A0709	Device Sensing Problem		
4394	2018	All other reportable incidents			CT257	Implantable cardiac pacemakers	F1905	Device Revision or Replacement	A07	Electrical /Electronic Property Problem	09/03/2018	
									A0708	Power Problem	09/03/2018	
4395	2017	All other reportable incidents	47270	Cardiac resynchronization therapy implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	F1905	Device Revision or Replacement	A07	Electrical /Electronic Property Problem	06/01/2017	
									A0708	Power Problem	06/01/2017	
4396	2017	All other reportable incidents			CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A1104	Computer Operating System Problem		

4397	2017	All other reportable incidents			CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A09	Output Problem		
									A0908	Incorrect, Inadequate or Imprecise Result or Readings		
4398	2018	All other reportable incidents	47270	Cardiac resynchronization therapy implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	F1905	Device Revision or Replacement	A07	Electrical /Electronic Property Problem	01/08/2018	
									A0708	Power Problem	01/08/2018	
4399	2017	Unanticipated serious deterioration in state of health			CT257	Implantable cardiac pacemakers	F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A07	Electrical /Electronic Property Problem		
									A0709	Device Sensing Problem		
4400	2017	TBD	47270	Cardiac resynchronization therapy implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A07	Electrical /Electronic Property Problem		
									A0708	Power Problem		
4401	2018	Death	35852	Single-chamber implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E2401	Insufficient Information	A07	Electrical /Electronic Property Problem	01/08/2018	
									A0709	Device Sensing Problem	01/08/2018	
4402	2017	TBD	35223	Endocardial/interventricular septal pacing lead	CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A26	Insufficient Device Problem Information	06/01/2017	
4403	2018	All other reportable incidents	47270	Cardiac resynchronization therapy implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	F1905	Device Revision or Replacement	A07	Electrical /Electronic Property Problem	25/07/2018	
									A0709	Device Sensing Problem	25/07/2018	
4404	2017	TBD	35223	Endocardial/interventricular septal pacing lead	CT1684	Cardiac pacemakers and associated devices	F14	Prolonged Episode of Care	A15	Activation, Positioning or Separation Problem		
									A150101	Activation Failure		
4405	2018	All other reportable incidents	35852	Single-chamber implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	F1905	Device Revision or Replacement	A1304	Telemetry Discrepancy		
									A26	Insufficient Device Problem Information		
4406	2017	All other reportable incidents			CT257	Implantable cardiac pacemakers	F1905	Device Revision or Replacement	A07	Electrical /Electronic Property Problem	17/05/2017	
									A0708	Power Problem	17/05/2017	
4407	2018	All other reportable incidents	47270	Cardiac resynchronization therapy implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A07	Electrical /Electronic Property Problem		
									A0709	Device Sensing Problem		
4408	2017	All other reportable incidents	47270	Cardiac resynchronization therapy implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A09	Output Problem		
									A0908	Incorrect, Inadequate or Imprecise Result or Readings		
4409	2017	All other reportable incidents	47270	Cardiac resynchronization therapy implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	F1905	Device Revision or Replacement	A07	Electrical /Electronic Property Problem	09/01/2017	
									A0708	Power Problem	09/01/2017	
4410	2017	All other reportable incidents	47270	Cardiac resynchronization therapy implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	F1905	Device Revision or Replacement	A07	Electrical /Electronic Property Problem	11/10/2017	
									A0708	Power Problem	11/10/2017	
4411	2018	Death			CT1684	Cardiac pacemakers and associated devices	F21	Unexpected Deterioration	A09	Output Problem		
									A0908	Incorrect, Inadequate or Imprecise Result or Readings		
4412	2017	All other reportable incidents			CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A09	Output Problem		
									A0907	No Device Output		
4413	2018	All other reportable incidents	47270	Cardiac resynchronization therapy implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	F1905	Device Revision or Replacement	A07	Electrical /Electronic Property Problem	20/07/2018	
									A0708	Power Problem	20/07/2018	
4414	2018	All other reportable incidents			CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A09	Output Problem		
									A0907	No Device Output		
4415	2017	TBD	47263	Cardiac resynchronization therapy implantable pacemaker	CT1684	Cardiac pacemakers and associated devices	F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A07	Electrical /Electronic Property Problem		
									A0709	Device Sensing Problem		
					CT1685	Cardiac pacemakers	F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A07	Electrical /Electronic Property Problem		
									A0709	Device Sensing Problem		
					CT257	Implantable cardiac pacemakers	F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A07	Electrical /Electronic Property Problem		
									A0709	Device Sensing Problem		
					CT2866	Implantable cardiac electrical stimulators	F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A07	Electrical /Electronic Property Problem		
									A0709	Device Sensing Problem		
4416	2018	All other reportable incidents			CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A09	Output Problem		
									A0907	No Device Output		
4417	2017	TBD	35223	Endocardial/interventricular septal pacing lead	CT1684	Cardiac pacemakers and associated devices			A26	Insufficient Device Problem Information		
4418	2018	All other reportable incidents	35223	Endocardial/interventricular septal pacing lead	CT1684	Cardiac pacemakers and associated devices	F1905	Device Revision or Replacement	A07	Electrical /Electronic Property Problem	01/03/2018	
									A0709	Device Sensing Problem	01/03/2018	
4419	2017	All other reportable incidents	47270	Cardiac resynchronization therapy implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	F1905	Device Revision or Replacement	A07	Electrical /Electronic Property Problem	11/12/2016	
									A0708	Power Problem	11/12/2016	
4420	2018	All other reportable incidents			CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A07	Electrical /Electronic Property Problem		
									A0708	Power Problem		
4421	2018	All other reportable incidents			CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A07	Electrical /Electronic Property Problem		
									A0709	Device Sensing Problem		
4422	2018	All other reportable incidents			CT257	Implantable cardiac pacemakers	F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A07	Electrical /Electronic Property Problem		
									A0708	Power Problem		
4423	2018	All other reportable incidents	47270	Cardiac resynchronization therapy implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A07	Electrical /Electronic Property Problem		
									A0708	Power Problem		
4424	2018	All other reportable incidents			CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A09	Output Problem		
									A0907	No Device Output		
4425	2017	All other reportable incidents	35223	Endocardial/interventricular septal pacing lead	CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A04	Material Integrity Problem		
									A0405	Degraded		
4426	2018	All other reportable incidents	35852	Single-chamber implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A09	Output Problem		
									A090404	Output Above Specifications		
4427	2017	All other reportable incidents	47270	Cardiac resynchronization therapy implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A11	Computer Software Problem		
									A1102	Application Program Problem		
4428	2018	All other reportable incidents	35852	Single-chamber implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A09	Output Problem		
									A090404	Output Above Specifications		
4429	2018	All other reportable incidents	35223	Endocardial/interventricular septal pacing lead	CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A05	Mechanical Problem		
									A0510	Retraction Problem		
4430	2017	TBD	35223	Endocardial/interventricular septal pacing lead	CT1684	Cardiac pacemakers and associated devices	E2401	Insufficient Information	A26	Insufficient Device Problem Information		
4431	2017	All other reportable incidents	35852	Single-chamber implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A07	Electrical /Electronic Property Problem		
									A0709	Device Sensing Problem		
4432	2017	All other reportable incidents	47270	Cardiac resynchronization therapy implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	F1905	Device Revision or Replacement	A24	No Apparent Device Problem	06/12/2016	
									A26	Insufficient Device Problem Information	06/12/2016	
4433	2017	All other reportable incidents	47270	Cardiac resynchronization therapy implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	F1905	Device Revision or Replacement	A07	Electrical /Electronic Property Problem	07/12/2016	
									A0708	Power Problem	07/12/2016	
4434	2018	All other reportable incidents			CT257	Implantable cardiac pacemakers	F1905	Device Revision or Replacement	A07	Electrical /Electronic Property Problem	11/05/2018	
									A0709	Device Sensing Problem	11/05/2018	
4435	2018	All other reportable incidents			CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A07	Electrical /Electronic Property Problem		
									A0709	Device Sensing Problem		
4436	2018	All other reportable incidents	47263	Cardiac resynchronization therapy implantable pacemaker	CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A1104	Computer Operating System Problem		
					CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A1104	Computer Operating System Problem		

					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A1104	Computer Operating System Problem		
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A1104	Computer Operating System Problem		
4437	2018	All other reportable incidents			CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A17	Compatibility Problem		
									A1702	Device-Device Incompatibility		
4438	2017	TBD			CT257	Implantable cardiac pacemakers	F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A16	Protective Measures Problem		
									A1602	Fail-Safe Problem		
4439	2018	All other reportable incidents	35223	Endocardial/interventricular septal pacing lead	CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A0104	Migration or Expulsion of Device		
									A15	Activation, Positioning or Separation Problem		
									A150102	Difficult or Delayed Activation		
									A150203	Difficult or Delayed Positioning		
4440	2017	TBD	35223	Endocardial/interventricular septal pacing lead	CT1684	Cardiac pacemakers and associated devices	F14	Prolonged Episode of Care	A26	Insufficient Device Problem Information		
4441	2017	All other reportable incidents	47270	Cardiac resynchronization therapy implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	F1905	Device Revision or Replacement	A07	Electrical /Electronic Property Problem	11/10/2017	
									A0708	Power Problem	11/10/2017	
4442	2018	All other reportable incidents	35852	Single-chamber implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E1906	Unspecified Infection	A01	Patient Device Interaction Problem		
									A17	Compatibility Problem		
4443	2017	All other reportable incidents	60789	Intracardiac pacemaker	CT130	Catheters and associated devices	F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A09	Output Problem		
									A0907	No Device Output		
									A13	Communication or Transmission Problem		
									A26	Insufficient Device Problem Information		
					CT145	Cardiovascular devices	F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A09	Output Problem		
									A0907	No Device Output		
									A13	Communication or Transmission Problem		
									A26	Insufficient Device Problem Information		
					CT1684	Cardiac pacemakers and associated devices	F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A09	Output Problem		
									A0907	No Device Output		
									A13	Communication or Transmission Problem		
									A26	Insufficient Device Problem Information		
					CT1685	Cardiac pacemakers	F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A09	Output Problem		
									A0907	No Device Output		
									A13	Communication or Transmission Problem		
									A26	Insufficient Device Problem Information		
					CT233	Surgical	F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A09	Output Problem		
									A0907	No Device Output		
									A13	Communication or Transmission Problem		
									A26	Insufficient Device Problem Information		
					CT2406	Implantable	F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A09	Output Problem		
									A0907	No Device Output		
									A13	Communication or Transmission Problem		
									A26	Insufficient Device Problem Information		
					CT2407	Active implantable	F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A09	Output Problem		
									A0907	No Device Output		
									A13	Communication or Transmission Problem		
									A26	Insufficient Device Problem Information		
					CT257	Implantable cardiac pacemakers	F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A09	Output Problem		
									A0907	No Device Output		
									A13	Communication or Transmission Problem		
									A26	Insufficient Device Problem Information		
					CT2727	Active implant devices/ancillaries	F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A09	Output Problem		
									A0907	No Device Output		
									A13	Communication or Transmission Problem		
									A26	Insufficient Device Problem Information		
					CT2750	Implantable cardiac electrical stimulators and associated devices	F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A09	Output Problem		
									A0907	No Device Output		
									A13	Communication or Transmission Problem		
									A26	Insufficient Device Problem Information		
					CT2818	Catheterization/Cannulation/Fluid path devices	F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A09	Output Problem		
									A0907	No Device Output		
									A13	Communication or Transmission Problem		
									A26	Insufficient Device Problem Information		
					CT2866	Implantable cardiac electrical stimulators	F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A09	Output Problem		
									A0907	No Device Output		
									A13	Communication or Transmission Problem		
									A26	Insufficient Device Problem Information		
					CT2985	Active	F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A09	Output Problem		
									A0907	No Device Output		
									A13	Communication or Transmission Problem		
									A26	Insufficient Device Problem Information		
					CT3001	Implantable metal	F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A09	Output Problem		
									A0907	No Device Output		
									A13	Communication or Transmission Problem		
									A26	Insufficient Device Problem Information		
					CT3002	Implantable synthetic polymer	F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A09	Output Problem		
									A0907	No Device Output		
									A13	Communication or Transmission Problem		
									A26	Insufficient Device Problem Information		
					CT3006	Implantable inorganic material	F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A09	Output Problem		
									A0907	No Device Output		
									A13	Communication or Transmission Problem		
									A26	Insufficient Device Problem Information		
					CT314	Electrically powered	F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A09	Output Problem		
									A0907	No Device Output		
									A13	Communication or Transmission Problem		
									A26	Insufficient Device Problem Information		
					CT321	Long-term surgical invasive	F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A09	Output Problem		
									A0907	No Device Output		
									A13	Communication or Transmission Problem		
									A26	Insufficient Device Problem Information		

					CT981	Single use	F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A09	Output Problem		
									A0907	No Device Output		
									A13	Communication or Transmission Problem		
									A26	Insufficient Device Problem Information		
					CT983	Surgical invasive	F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A09	Output Problem		
									A0907	No Device Output		
									A13	Communication or Transmission Problem		
									A26	Insufficient Device Problem Information		
4444	2017	All other reportable incidents	35852	Single-chamber implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A07	Electrical /Electronic Property Problem		
									A0709	Device Sensing Problem		
4445	2017	All other reportable incidents			CT257	Implantable cardiac pacemakers	F1905	Device Revision or Replacement	A11	Computer Software Problem	04/01/2017	
									A1102	Application Program Problem	04/01/2017	
4446	2017	All other reportable incidents			CT257	Implantable cardiac pacemakers			A09	Output Problem		
									A0907	No Device Output		
4447	2017	All other reportable incidents	47270	Cardiac resynchronization therapy implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	F1905	Device Revision or Replacement	A07	Electrical /Electronic Property Problem	16/12/2016	
									A0708	Power Problem	16/12/2016	
4448	2017	Death	47270	Cardiac resynchronization therapy implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A09	Output Problem		
									A0907	No Device Output		
4449	2017	All other reportable incidents	47263	Cardiac resynchronization therapy implantable pacemaker	CT1684	Cardiac pacemakers and associated devices	F1905	Device Revision or Replacement	A07	Electrical /Electronic Property Problem	15/12/2017	
									A0709	Device Sensing Problem	15/12/2017	
					CT1685	Cardiac pacemakers	F1905	Device Revision or Replacement	A07	Electrical /Electronic Property Problem	15/12/2017	
									A0709	Device Sensing Problem	15/12/2017	
					CT257	Implantable cardiac pacemakers	F1905	Device Revision or Replacement	A07	Electrical /Electronic Property Problem	15/12/2017	
									A0709	Device Sensing Problem	15/12/2017	
					CT2866	Implantable cardiac electrical stimulators	F1905	Device Revision or Replacement	A07	Electrical /Electronic Property Problem	15/12/2017	
									A0709	Device Sensing Problem	15/12/2017	
4450	2017	All other reportable incidents	47263	Cardiac resynchronization therapy implantable pacemaker	CT1684	Cardiac pacemakers and associated devices	F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A07	Electrical /Electronic Property Problem		
									A0709	Device Sensing Problem		
					CT1685	Cardiac pacemakers	F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A07	Electrical /Electronic Property Problem		
									A0709	Device Sensing Problem		
					CT257	Implantable cardiac pacemakers	F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A07	Electrical /Electronic Property Problem		
									A0709	Device Sensing Problem		
					CT2866	Implantable cardiac electrical stimulators	F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A07	Electrical /Electronic Property Problem		
									A0709	Device Sensing Problem		
4451	2017	All other reportable incidents			CT1684	Cardiac pacemakers and associated devices	F05	Delay to Treatment/ Therapy	A07	Electrical /Electronic Property Problem		
									A0708	Power Problem		
4452	2017	All other reportable incidents	35223	Endocardial/interventricular septal pacing lead	CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A15	Activation, Positioning or Separation Problem		
									A150101	Activation Failure		
4453	2018	All other reportable incidents	47270	Cardiac resynchronization therapy implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A07	Electrical /Electronic Property Problem		
									A0708	Power Problem		
4454	2018	Death	47270	Cardiac resynchronization therapy implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E2401	Insufficient Information	A1304	Telemetry Discrepancy		
									A26	Insufficient Device Problem Information		
4455	2017	All other reportable incidents	47263	Cardiac resynchronization therapy implantable pacemaker	CT1684	Cardiac pacemakers and associated devices	F1905	Device Revision or Replacement	A11	Computer Software Problem		
									A1102	Application Program Problem		
					CT1685	Cardiac pacemakers	F1905	Device Revision or Replacement	A11	Computer Software Problem		
									A1102	Application Program Problem		
					CT257	Implantable cardiac pacemakers	F1905	Device Revision or Replacement	A11	Computer Software Problem		
									A1102	Application Program Problem		
					CT2866	Implantable cardiac electrical stimulators	F1905	Device Revision or Replacement	A11	Computer Software Problem		
									A1102	Application Program Problem		
4456	2017	TBD			CT1684	Cardiac pacemakers and associated devices	E2104	Electric Shock	A05	Mechanical Problem		
									A051201	Device Dislodged or Dislocated		
									A1104	Computer Operating System Problem		
4457	2018	All other reportable incidents			CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A09	Output Problem		
									A0907	No Device Output		
4458	2018	All other reportable incidents	35852	Single-chamber implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A07	Electrical /Electronic Property Problem		
									A0709	Device Sensing Problem		
									A09	Output Problem		
									A090404	Output Above Specifications		
4459	2017	All other reportable incidents	47270	Cardiac resynchronization therapy implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	F1905	Device Revision or Replacement	A07	Electrical /Electronic Property Problem	06/01/2017	
									A0708	Power Problem	06/01/2017	
4460	2017	All other reportable incidents	35223	Endocardial/interventricular septal pacing lead	CT1684	Cardiac pacemakers and associated devices	F1905	Device Revision or Replacement	A05	Mechanical Problem		
									A051201	Device Dislodged or Dislocated		
									A15	Activation, Positioning or Separation Problem		
									A150203	Difficult or Delayed Positioning		
4461	2018	All other reportable incidents			CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A07	Electrical /Electronic Property Problem		
									A0709	Device Sensing Problem		
4462	2017	TBD	35223	Endocardial/interventricular septal pacing lead	CT1684	Cardiac pacemakers and associated devices	F1905	Device Revision or Replacement	A0104	Migration or Expulsion of Device		
4463	2017	All other reportable incidents	47270	Cardiac resynchronization therapy implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	F1905	Device Revision or Replacement	A07	Electrical /Electronic Property Problem	21/11/2016	
									A0708	Power Problem	21/11/2016	
4464	2018	All other reportable incidents	35852	Single-chamber implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A07	Electrical /Electronic Property Problem		
									A0709	Device Sensing Problem		
4465	2017	All other reportable incidents	47263	Cardiac resynchronization therapy implantable pacemaker	CT1684	Cardiac pacemakers and associated devices	F1905	Device Revision or Replacement	A07	Electrical /Electronic Property Problem		
									A0708	Power Problem		
					CT1685	Cardiac pacemakers	F1905	Device Revision or Replacement	A07	Electrical /Electronic Property Problem		
									A0708	Power Problem		
					CT257	Implantable cardiac pacemakers	F1905	Device Revision or Replacement	A07	Electrical /Electronic Property Problem		
									A0708	Power Problem		
					CT2866	Implantable cardiac electrical stimulators	F1905	Device Revision or Replacement	A07	Electrical /Electronic Property Problem		
									A0708	Power Problem		
4466	2017	TBD	35223	Endocardial/interventricular septal pacing lead	CT1684	Cardiac pacemakers and associated devices			A26	Insufficient Device Problem Information		
4467	2017	All other reportable incidents	47270	Cardiac resynchronization therapy implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A07	Electrical /Electronic Property Problem	18/01/2017	
									A0708	Power Problem	18/01/2017	
4468	2017	All other reportable incidents	47270	Cardiac resynchronization therapy implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	F1905	Device Revision or Replacement	A07	Electrical /Electronic Property Problem	10/01/2017	
									A0708	Power Problem	10/01/2017	

4469	2017	TBD	35223	Endocardial/interventricular septal pacing lead	CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A0104	Migration or Expulsion of Device		
4470	2017	TBD			CT1684	Cardiac pacemakers and associated devices	F05	Delay to Treatment/ Therapy	A26	Insufficient Device Problem Information		
4471	2017	TBD	47263	Cardiac resynchronization therapy implantable pacemaker	CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A07	Electrical /Electronic Property Problem		
									A0709	Device Sensing Problem		
					CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A07	Electrical /Electronic Property Problem		
									A0709	Device Sensing Problem		
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A07	Electrical /Electronic Property Problem		
									A0709	Device Sensing Problem		
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A07	Electrical /Electronic Property Problem		
									A0709	Device Sensing Problem		
4472	2017	TBD	35223	Endocardial/interventricular septal pacing lead	CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A15	Activation, Positioning or Separation Problem		
									A150203	Difficult or Delayed Positioning		
4473	2017	TBD	47270	Cardiac resynchronization therapy implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A07	Electrical /Electronic Property Problem		
									A0709	Device Sensing Problem		
4474	2017	All other reportable incidents	35223	Endocardial/interventricular septal pacing lead	CT1684	Cardiac pacemakers and associated devices	E2401	Insufficient Information	A26	Insufficient Device Problem Information		
4475	2017	All other reportable incidents			CT257	Implantable cardiac pacemakers	E2401	Insufficient Information	A26	Insufficient Device Problem Information		
4476	2017	TBD	47263	Cardiac resynchronization therapy implantable pacemaker	CT1684	Cardiac pacemakers and associated devices	F1905	Device Revision or Replacement	A13	Communication or Transmission Problem		
									A26	Insufficient Device Problem Information		
					CT1685	Cardiac pacemakers	F1905	Device Revision or Replacement	A13	Communication or Transmission Problem		
									A26	Insufficient Device Problem Information		
					CT257	Implantable cardiac pacemakers	F1905	Device Revision or Replacement	A13	Communication or Transmission Problem		
									A26	Insufficient Device Problem Information		
					CT2866	Implantable cardiac electrical stimulators	F1905	Device Revision or Replacement	A13	Communication or Transmission Problem		
									A26	Insufficient Device Problem Information		
4477	2018	All other reportable incidents	35223	Endocardial/interventricular septal pacing lead	CT1684	Cardiac pacemakers and associated devices	F1905	Device Revision or Replacement	A0104	Migration or Expulsion of Device		
4478	2018	All other reportable incidents	47270	Cardiac resynchronization therapy implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A16	Protective Measures Problem		
									A1602	Fail-Safe Problem		
4479	2018	All other reportable incidents	47270	Cardiac resynchronization therapy implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A26	Insufficient Device Problem Information		
4480	2017	All other reportable incidents	47270	Cardiac resynchronization therapy implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A07	Electrical /Electronic Property Problem		
									A0709	Device Sensing Problem		
4481	2018	All other reportable incidents			CT257	Implantable cardiac pacemakers	F1905	Device Revision or Replacement	A13	Communication or Transmission Problem	04/09/2018	
									A16	Protective Measures Problem	04/09/2018	
									A1602	Fail-Safe Problem	04/09/2018	
									A26	Insufficient Device Problem Information	04/09/2018	
4482	2017	All other reportable incidents	47270	Cardiac resynchronization therapy implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	F1905	Device Revision or Replacement	A07	Electrical /Electronic Property Problem		
									A0708	Power Problem		
4483	2017	All other reportable incidents	47270	Cardiac resynchronization therapy implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	F1905	Device Revision or Replacement	A07	Electrical /Electronic Property Problem		
									A0708	Power Problem		
4484	2018	All other reportable incidents	35852	Single-chamber implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	F1905	Device Revision or Replacement	A07	Electrical /Electronic Property Problem	23/03/2018	
									A0708	Power Problem	23/03/2018	
4485	2018	All other reportable incidents	47270	Cardiac resynchronization therapy implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A07	Electrical /Electronic Property Problem		
									A0709	Device Sensing Problem		
4486	2017	TBD	47270	Cardiac resynchronization therapy implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	F1905	Device Revision or Replacement	A09	Output Problem		
									A0907	No Device Output		
4487	2017	TBD	47270	Cardiac resynchronization therapy implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	F1905	Device Revision or Replacement	A07	Electrical /Electronic Property Problem		
									A0708	Power Problem		
4488	2018	All other reportable incidents	47270	Cardiac resynchronization therapy implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A07	Electrical /Electronic Property Problem		
									A0708	Power Problem		
4489	2018	All other reportable incidents			CT257	Implantable cardiac pacemakers	F1905	Device Revision or Replacement	A07	Electrical /Electronic Property Problem		
									A0708	Power Problem		
4490	2017	All other reportable incidents	35223	Endocardial/interventricular septal pacing lead	CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A05	Mechanical Problem	31/10/2017	
									A0510	Retraction Problem	31/10/2017	
4491	2017	TBD	47263	Cardiac resynchronization therapy implantable pacemaker	CT1684	Cardiac pacemakers and associated devices	E2007	Fall	A07	Electrical /Electronic Property Problem		
									A0709	Device Sensing Problem		
					CT1685	Cardiac pacemakers	E2007	Fall	A07	Electrical /Electronic Property Problem		
									A0709	Device Sensing Problem		
					CT257	Implantable cardiac pacemakers	E2007	Fall	A07	Electrical /Electronic Property Problem		
									A0709	Device Sensing Problem		
					CT2866	Implantable cardiac electrical stimulators	E2007	Fall	A07	Electrical /Electronic Property Problem		
									A0709	Device Sensing Problem		
4492	2017	All other reportable incidents	35223	Endocardial/interventricular septal pacing lead	CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A15	Activation, Positioning or Separation Problem		
									A150203	Difficult or Delayed Positioning		
4493	2017	All other reportable incidents	47270	Cardiac resynchronization therapy implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A07	Electrical /Electronic Property Problem		
									A0708	Power Problem		
4494	2017	All other reportable incidents	47263	Cardiac resynchronization therapy implantable pacemaker	CT1684	Cardiac pacemakers and associated devices	F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A07	Electrical /Electronic Property Problem		
									A0709	Device Sensing Problem		
					CT1685	Cardiac pacemakers	F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A07	Electrical /Electronic Property Problem		
									A0709	Device Sensing Problem		
					CT257	Implantable cardiac pacemakers	F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A07	Electrical /Electronic Property Problem		
									A0709	Device Sensing Problem		
					CT2866	Implantable cardiac electrical stimulators	F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A07	Electrical /Electronic Property Problem		
									A0709	Device Sensing Problem		
4495	2017	All other reportable incidents	35852	Single-chamber implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A07	Electrical /Electronic Property Problem		
									A0709	Device Sensing Problem		
4496	2018	All other reportable incidents	35852	Single-chamber implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A07	Electrical /Electronic Property Problem		
									A0709	Device Sensing Problem		
4497	2018	All other reportable incidents			CT257	Implantable cardiac pacemakers	F1905	Device Revision or Replacement	A07	Electrical /Electronic Property Problem	21/08/2018	
									A0708	Power Problem	21/08/2018	
4498	2017	All other reportable incidents	47270	Cardiac resynchronization therapy implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	F1905	Device Revision or Replacement	A24	No Apparent Device Problem	19/01/2017	
									A26	Insufficient Device Problem Information	19/01/2017	
4499	2017	Death	47270	Cardiac resynchronization therapy implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E2401	Insufficient Information	A26	Insufficient Device Problem Information		
4500	2018	All other reportable incidents	35852	Single-chamber implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A09	Output Problem	21/08/2018	
									A090404	Output Above Specifications	21/08/2018	
							F1905	Device Revision or Replacement	A09	Output Problem	21/08/2018	

									A090404	Output Above Specifications	21/08/2018	
4501	2018	All other reportable incidents	35852	Single-chamber implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A07	Electrical /Electronic Property Problem		
									A0709	Device Sensing Problem		
4502	2017	TBD	35223	Endocardial/interventricular septal pacing lead	CT1684	Cardiac pacemakers and associated devices	F14	Prolonged Episode of Care	A07	Electrical /Electronic Property Problem		
									A0709	Device Sensing Problem		
4503	2017	TBD	47270	Cardiac resynchronization therapy implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	F1905	Device Revision or Replacement	A07	Electrical /Electronic Property Problem		
									A0708	Power Problem		
4504	2017	All other reportable incidents	47270	Cardiac resynchronization therapy implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	F1905	Device Revision or Replacement	A07	Electrical /Electronic Property Problem		
									A0708	Power Problem		
4505	2017	TBD	35223	Endocardial/interventricular septal pacing lead	CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A15	Activation, Positioning or Separation Problem		
									A150203	Difficult or Delayed Positioning		
4506	2017	TBD	47270	Cardiac resynchronization therapy implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A07	Electrical /Electronic Property Problem		
									A0708	Power Problem		
4507	2017	All other reportable incidents	35852	Single-chamber implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E1906	Unspecified Infection	A01	Patient Device Interaction Problem		
									A17	Compatibility Problem		
4508	2017	TBD	35852	Single-chamber implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A07	Electrical /Electronic Property Problem		
									A0709	Device Sensing Problem		
4509	2017	All other reportable incidents	35223	Endocardial/interventricular septal pacing lead	CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A15	Activation, Positioning or Separation Problem		
									A150203	Difficult or Delayed Positioning		
4510	2017	All other reportable incidents	35223	Endocardial/interventricular septal pacing lead	CT1684	Cardiac pacemakers and associated devices	E2401	Insufficient Information	A26	Insufficient Device Problem Information		
4511	2017	All other reportable incidents			CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A07	Electrical /Electronic Property Problem	19/10/2016	
									A0708	Power Problem	19/10/2016	
4512	2017	All other reportable incidents	47270	Cardiac resynchronization therapy implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	F1905	Device Revision or Replacement	A07	Electrical /Electronic Property Problem	19/01/2017	
									A0708	Power Problem	19/01/2017	
4513	2017	All other reportable incidents	47270	Cardiac resynchronization therapy implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	F1905	Device Revision or Replacement	A07	Electrical /Electronic Property Problem	19/01/2017	
									A0708	Power Problem	19/01/2017	
4514	2018	All other reportable incidents			CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A07	Electrical /Electronic Property Problem		
									A0709	Device Sensing Problem		
4515	2017	All other reportable incidents	47270	Cardiac resynchronization therapy implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	F1905	Device Revision or Replacement	A07	Electrical /Electronic Property Problem		
									A0708	Power Problem		
4516	2017	All other reportable incidents	47270	Cardiac resynchronization therapy implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A07	Electrical /Electronic Property Problem		
									A0709	Device Sensing Problem		
4517	2017	All other reportable incidents	47270	Cardiac resynchronization therapy implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	F1905	Device Revision or Replacement	A07	Electrical /Electronic Property Problem	20/01/2017	
									A0708	Power Problem	20/01/2017	
4518	2017	All other reportable incidents	35852	Single-chamber implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A07	Electrical /Electronic Property Problem		
									A0709	Device Sensing Problem		
									A09	Output Problem		
									A090404	Output Above Specifications		
4519	2017	All other reportable incidents			CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A07	Electrical /Electronic Property Problem		
									A0709	Device Sensing Problem		
4520	2017	All other reportable incidents	35852	Single-chamber implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A1104	Computer Operating System Problem		
4521	2017	TBD	47270	Cardiac resynchronization therapy implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A26	Insufficient Device Problem Information		
4522	2017	TBD	47270	Cardiac resynchronization therapy implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A07	Electrical /Electronic Property Problem		
									A0708	Power Problem		
4523	2017	All other reportable incidents			CT257	Implantable cardiac pacemakers	F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A07	Electrical /Electronic Property Problem		
									A0709	Device Sensing Problem		
4524	2017	All other reportable incidents			CT257	Implantable cardiac pacemakers	F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A07	Electrical /Electronic Property Problem		
									A0709	Device Sensing Problem		
4525	2017	All other reportable incidents			CT257	Implantable cardiac pacemakers	F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A07	Electrical /Electronic Property Problem		
									A0709	Device Sensing Problem		
4526	2017	TBD			CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A07	Electrical /Electronic Property Problem		
									A0709	Device Sensing Problem		
4527	2018	All other reportable incidents	35223	Endocardial/interventricular septal pacing lead	CT1684	Cardiac pacemakers and associated devices	F1905	Device Revision or Replacement	A07	Electrical /Electronic Property Problem	29/08/2018	
									A0709	Device Sensing Problem	29/08/2018	
4528	2017	All other reportable incidents	47270	Cardiac resynchronization therapy implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A07	Electrical /Electronic Property Problem		
									A0708	Power Problem		
4529	2017	All other reportable incidents			CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A16	Protective Measures Problem		
									A1602	Fail-Safe Problem		
4530	2017	All other reportable incidents			CT257	Implantable cardiac pacemakers	F1905	Device Revision or Replacement	A07	Electrical /Electronic Property Problem	23/10/2017	
									A0708	Power Problem	23/10/2017	
4531	2017	All other reportable incidents			CT1684	Cardiac pacemakers and associated devices	F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A11	Computer Software Problem		
									A1102	Application Program Problem		
4532	2018	All other reportable incidents	35223	Endocardial/interventricular septal pacing lead	CT1684	Cardiac pacemakers and associated devices	F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A0104	Migration or Expulsion of Device		
4533	2018	All other reportable incidents	35223	Endocardial/interventricular septal pacing lead	CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A07	Electrical /Electronic Property Problem		
									A0709	Device Sensing Problem		
4534	2017	All other reportable incidents	35223	Endocardial/interventricular septal pacing lead	CT1684	Cardiac pacemakers and associated devices	F1905	Device Revision or Replacement	A04	Material Integrity Problem	01/06/2017	
									A0405	Degraded	01/06/2017	
									A07	Electrical /Electronic Property Problem	01/06/2017	
									A0709	Device Sensing Problem	01/06/2017	
4535	2017	All other reportable incidents	35223	Endocardial/interventricular septal pacing lead	CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A15	Activation, Positioning or Separation Problem		
									A150101	Activation Failure		
4536	2017	TBD	35223	Endocardial/interventricular septal pacing lead	CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A15	Activation, Positioning or Separation Problem		
									A150203	Difficult or Delayed Positioning		
4537	2018	All other reportable incidents	47270	Cardiac resynchronization therapy implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A07	Electrical /Electronic Property Problem		
									A0709	Device Sensing Problem		
4538	2017	TBD	35223	Endocardial/interventricular septal pacing lead	CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A04	Material Integrity Problem		
									A0405	Degraded		
4539	2017	All other reportable incidents			CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A05	Mechanical Problem		
									A0506	Mechanical Jam		
4540	2017	TBD	35223	Endocardial/interventricular septal pacing lead	CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A15	Activation, Positioning or Separation Problem		
									A150203	Difficult or Delayed Positioning		
4541	2017	All other reportable incidents	47270	Cardiac resynchronization therapy implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	F1905	Device Revision or Replacement	A07	Electrical /Electronic Property Problem	19/01/2017	
									A0708	Power Problem	19/01/2017	
4542	2018	All other reportable incidents	35223	Endocardial/interventricular septal pacing lead	CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A0104	Migration or Expulsion of Device		
4543	2017	TBD	47270	Cardiac resynchronization therapy implantable	CT2866	Implantable cardiac electrical stimulators	F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A07	Electrical /Electronic Property Problem		

				defibrillator					A0708	Power Problem		
4544	2017	TBD	47270	Cardiac resynchronization therapy implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A07	Electrical /Electronic Property Problem		
									A0708	Power Problem		
4545	2018	All other reportable incidents			CT257	Implantable cardiac pacemakers	F1905	Device Revision or Replacement	A07	Electrical /Electronic Property Problem	16/03/2018	
									A0708	Power Problem	16/03/2018	
4546	2017	All other reportable incidents	47263	Cardiac resynchronization therapy implantable pacemaker	CT1684	Cardiac pacemakers and associated devices	F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A07	Electrical /Electronic Property Problem		
									A0709	Device Sensing Problem		
					CT1685	Cardiac pacemakers	F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A07	Electrical /Electronic Property Problem		
									A0709	Device Sensing Problem		
					CT257	Implantable cardiac pacemakers	F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A07	Electrical /Electronic Property Problem		
									A0709	Device Sensing Problem		
					CT2866	Implantable cardiac electrical stimulators	F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A07	Electrical /Electronic Property Problem		
									A0709	Device Sensing Problem		
4547	2017	TBD	47270	Cardiac resynchronization therapy implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A07	Electrical /Electronic Property Problem		
									A0709	Device Sensing Problem		
4548	2017	TBD	47270	Cardiac resynchronization therapy implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A07	Electrical /Electronic Property Problem		
									A0708	Power Problem		
4549	2018	All other reportable incidents	47270	Cardiac resynchronization therapy implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A07	Electrical /Electronic Property Problem		
									A0709	Device Sensing Problem		
4550	2018	All other reportable incidents	47263	Cardiac resynchronization therapy implantable pacemaker	CT1684	Cardiac pacemakers and associated devices	F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A07	Electrical /Electronic Property Problem		
									A0709	Device Sensing Problem		
					CT1685	Cardiac pacemakers	F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A07	Electrical /Electronic Property Problem		
									A0709	Device Sensing Problem		
					CT257	Implantable cardiac pacemakers	F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A07	Electrical /Electronic Property Problem		
									A0709	Device Sensing Problem		
					CT2866	Implantable cardiac electrical stimulators	F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A07	Electrical /Electronic Property Problem		
									A0709	Device Sensing Problem		
4551	2018	All other reportable incidents	35223	Endocardial/interventricular septal pacing lead	CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A09	Output Problem		
									A0907	No Device Output		
4552	2017	TBD			CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A16	Protective Measures Problem		
									A1602	Fail-Safe Problem		
4553	2017	TBD	47270	Cardiac resynchronization therapy implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	F21	Unexpected Deterioration	A07	Electrical /Electronic Property Problem		
									A0708	Power Problem		
4554	2017	All other reportable incidents	47270	Cardiac resynchronization therapy implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	F1905	Device Revision or Replacement	A24	No Apparent Device Problem	23/01/2017	
									A26	Insufficient Device Problem Information	23/01/2017	
4555	2017	All other reportable incidents	47270	Cardiac resynchronization therapy implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	F1905	Device Revision or Replacement	A07	Electrical /Electronic Property Problem	13/01/2017	
									A0708	Power Problem	13/01/2017	
4556	2017	All other reportable incidents	47270	Cardiac resynchronization therapy implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	F1905	Device Revision or Replacement	A07	Electrical /Electronic Property Problem	24/01/2017	
									A0708	Power Problem	24/01/2017	
4557	2017	All other reportable incidents	47270	Cardiac resynchronization therapy implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	F1905	Device Revision or Replacement	A07	Electrical /Electronic Property Problem	24/01/2017	
									A0708	Power Problem	24/01/2017	
4558	2017	TBD	47270	Cardiac resynchronization therapy implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A1304	Telemetry Discrepancy		
									A26	Insufficient Device Problem Information		
4559	2017	TBD	47270	Cardiac resynchronization therapy implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	F1905	Device Revision or Replacement	A07	Electrical /Electronic Property Problem		
									A0708	Power Problem		
4560	2017	TBD	47270	Cardiac resynchronization therapy implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	F1905	Device Revision or Replacement	A07	Electrical /Electronic Property Problem		
									A0708	Power Problem		
4561	2017	TBD			CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A07	Electrical /Electronic Property Problem		
									A0709	Device Sensing Problem		
4562	2017	All other reportable incidents			CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A1104	Computer Operating System Problem		
4563	2018	All other reportable incidents			CT257	Implantable cardiac pacemakers	F1905	Device Revision or Replacement	A07	Electrical /Electronic Property Problem	28/08/2018	
									A0708	Power Problem	28/08/2018	
4564	2017	TBD	35852	Single-chamber implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A07	Electrical /Electronic Property Problem		
									A0709	Device Sensing Problem		
4565	2017	TBD	35223	Endocardial/interventricular septal pacing lead	CT1684	Cardiac pacemakers and associated devices	E2401	Insufficient Information	A26	Insufficient Device Problem Information		
4566	2017	All other reportable incidents	47270	Cardiac resynchronization therapy implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A07	Electrical /Electronic Property Problem		
									A0709	Device Sensing Problem		
4567	2017	TBD	47270	Cardiac resynchronization therapy implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A07	Electrical /Electronic Property Problem		
									A0709	Device Sensing Problem		
4568	2017	TBD	47270	Cardiac resynchronization therapy implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A1104	Computer Operating System Problem		
4569	2017	TBD	35223	Endocardial/interventricular septal pacing lead	CT1684	Cardiac pacemakers and associated devices	F21	Unexpected Deterioration	A26	Insufficient Device Problem Information		
4570	2018	All other reportable incidents	47263	Cardiac resynchronization therapy implantable pacemaker	CT1684	Cardiac pacemakers and associated devices	F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A07	Electrical /Electronic Property Problem	06/04/2018	
									A0709	Device Sensing Problem	06/04/2018	
							F14	Prolonged Episode of Care	A07	Electrical /Electronic Property Problem	06/04/2018	
									A0709	Device Sensing Problem	06/04/2018	
							F1905	Device Revision or Replacement	A07	Electrical /Electronic Property Problem	06/04/2018	
									A0709	Device Sensing Problem	06/04/2018	
					CT1685	Cardiac pacemakers	F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A07	Electrical /Electronic Property Problem	06/04/2018	
									A0709	Device Sensing Problem	06/04/2018	
							F14	Prolonged Episode of Care	A07	Electrical /Electronic Property Problem	06/04/2018	
									A0709	Device Sensing Problem	06/04/2018	
							F1905	Device Revision or Replacement	A07	Electrical /Electronic Property Problem	06/04/2018	
									A0709	Device Sensing Problem	06/04/2018	
					CT257	Implantable cardiac pacemakers	F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A07	Electrical /Electronic Property Problem	06/04/2018	
									A0709	Device Sensing Problem	06/04/2018	
							F14	Prolonged Episode of Care	A07	Electrical /Electronic Property Problem	06/04/2018	
									A0709	Device Sensing Problem	06/04/2018	
							F1905	Device Revision or Replacement	A07	Electrical /Electronic Property Problem	06/04/2018	
									A0709	Device Sensing Problem	06/04/2018	
					CT2866	Implantable cardiac electrical stimulators	F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A07	Electrical /Electronic Property Problem	06/04/2018	
									A0709	Device Sensing Problem	06/04/2018	
							F14	Prolonged Episode of Care	A07	Electrical /Electronic Property Problem	06/04/2018	
									A0709	Device Sensing Problem	06/04/2018	
							F1905	Device Revision or Replacement	A07	Electrical /Electronic Property Problem	06/04/2018	
									A0709	Device Sensing Problem	06/04/2018	

4571	2018	All other reportable incidents	35852	Single-chamber implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A07	Electrical /Electronic Property Problem		
									A0709	Device Sensing Problem		
4572	2017	TBD			CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A07	Electrical /Electronic Property Problem		
									A0709	Device Sensing Problem		
4573	2018	All other reportable incidents			CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A07	Electrical /Electronic Property Problem		
									A0709	Device Sensing Problem		
4574	2017	All other reportable incidents			CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A07	Electrical /Electronic Property Problem		
									A0708	Power Problem		
4575	2017	All other reportable incidents	47270	Cardiac resynchronization therapy implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A11	Computer Software Problem		
									A1102	Application Program Problem		
4576	2018	All other reportable incidents	35852	Single-chamber implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A07	Electrical /Electronic Property Problem		
									A0709	Device Sensing Problem		
									A26	Insufficient Device Problem Information		
4577	2018	All other reportable incidents	35852	Single-chamber implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A07	Electrical /Electronic Property Problem	26/10/2017	
									A0709	Device Sensing Problem	26/10/2017	
4578	2018	All other reportable incidents	35852	Single-chamber implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A07	Electrical /Electronic Property Problem		
									A0709	Device Sensing Problem		
									A09	Output Problem		
									A090404	Output Above Specifications		
4579	2018	All other reportable incidents	47270	Cardiac resynchronization therapy implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A07	Electrical /Electronic Property Problem		
									A0709	Device Sensing Problem		
									A09	Output Problem		
									A090404	Output Above Specifications		
4580	2017	TBD			CT257	Implantable cardiac pacemakers	F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A07	Electrical /Electronic Property Problem		
									A0709	Device Sensing Problem		
4581	2018	All other reportable incidents	35223	Endocardial/interventricular septal pacing lead	CT1684	Cardiac pacemakers and associated devices	F1905	Device Revision or Replacement	A0104	Migration or Expulsion of Device	21/03/2018	
4582	2018	All other reportable incidents	35852	Single-chamber implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A07	Electrical /Electronic Property Problem		
									A0709	Device Sensing Problem		
4583	2017	TBD			CT257	Implantable cardiac pacemakers	E2401	Insufficient Information	A26	Insufficient Device Problem Information		
4584	2017	All other reportable incidents	47270	Cardiac resynchronization therapy implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A07	Electrical /Electronic Property Problem		
									A0708	Power Problem		
4585	2017	TBD	47270	Cardiac resynchronization therapy implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	F1905	Device Revision or Replacement	A07	Electrical /Electronic Property Problem	22/06/2017	
									A0708	Power Problem	22/06/2017	
4586	2017	TBD	47270	Cardiac resynchronization therapy implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A07	Electrical /Electronic Property Problem		
									A0708	Power Problem		
4587	2017	TBD	35223	Endocardial/interventricular septal pacing lead	CT1684	Cardiac pacemakers and associated devices	E1906	Unspecified Infection	A07	Electrical /Electronic Property Problem		
									A0709	Device Sensing Problem		
4588	2017	TBD	35223	Endocardial/interventricular septal pacing lead	CT1684	Cardiac pacemakers and associated devices	E2401	Insufficient Information	A26	Insufficient Device Problem Information		
4589	2018	All other reportable incidents	47270	Cardiac resynchronization therapy implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A07	Electrical /Electronic Property Problem		
									A0709	Device Sensing Problem		
4590	2018	All other reportable incidents			CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A07	Electrical /Electronic Property Problem		
									A0709	Device Sensing Problem		
									A1104	Computer Operating System Problem		
4591	2017	All other reportable incidents			CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A1104	Computer Operating System Problem		
4592	2017	All other reportable incidents	47270	Cardiac resynchronization therapy implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	F1905	Device Revision or Replacement	A07	Electrical /Electronic Property Problem	05/01/2017	
									A0708	Power Problem	05/01/2017	
4593	2018	All other reportable incidents	35852	Single-chamber implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A09	Output Problem		
									A090404	Output Above Specifications		
4594	2018	All other reportable incidents	35852	Single-chamber implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A07	Electrical /Electronic Property Problem		
									A0709	Device Sensing Problem		
									A09	Output Problem		
									A090404	Output Above Specifications		
4595	2017	TBD	35223	Endocardial/interventricular septal pacing lead	CT1684	Cardiac pacemakers and associated devices	F1905	Device Revision or Replacement	A07	Electrical /Electronic Property Problem		
									A0709	Device Sensing Problem		
4596	2018	All other reportable incidents	35223	Endocardial/interventricular septal pacing lead	CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A05	Mechanical Problem		
									A0510	Retraction Problem		
4597	2017	TBD	35223	Endocardial/interventricular septal pacing lead	CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A07	Electrical /Electronic Property Problem		
									A0709	Device Sensing Problem		
4598	2018	All other reportable incidents	47270	Cardiac resynchronization therapy implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A07	Electrical /Electronic Property Problem	13/04/2018	
									A0708	Power Problem	13/04/2018	
4599	2017	TBD	35223	Endocardial/interventricular septal pacing lead	CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A15	Activation, Positioning or Separation Problem		
									A150101	Activation Failure		
4600	2018	All other reportable incidents	47270	Cardiac resynchronization therapy implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A07	Electrical /Electronic Property Problem		
									A0709	Device Sensing Problem		
4601	2017	TBD			CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A07	Electrical /Electronic Property Problem		
									A0709	Device Sensing Problem		
4602	2017	TBD			CT257	Implantable cardiac pacemakers	E2401	Insufficient Information	A26	Insufficient Device Problem Information		
4603	2018	All other reportable incidents			CT257	Implantable cardiac pacemakers	F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A09	Output Problem		
									A090404	Output Above Specifications		
							F1905	Device Revision or Replacement	A09	Output Problem		
									A090404	Output Above Specifications		
4604	2018	All other reportable incidents	47263	Cardiac resynchronization therapy implantable pacemaker	CT1684	Cardiac pacemakers and associated devices	F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A07	Electrical /Electronic Property Problem		
									A0709	Device Sensing Problem		
					CT1685	Cardiac pacemakers	F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A07	Electrical /Electronic Property Problem		
									A0709	Device Sensing Problem		
					CT257	Implantable cardiac pacemakers	F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A07	Electrical /Electronic Property Problem		
									A0709	Device Sensing Problem		
					CT2866	Implantable cardiac electrical stimulators	F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A07	Electrical /Electronic Property Problem		
									A0709	Device Sensing Problem		
4605	2018	All other reportable incidents	35852	Single-chamber implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A07	Electrical /Electronic Property Problem		
									A0709	Device Sensing Problem		
4606	2017	All other reportable incidents	35852	Single-chamber implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E2109	Implant Pain	A01	Patient Device Interaction Problem		
									A17	Compatibility Problem		
							F1905	Device Revision or Replacement	A01	Patient Device Interaction Problem		
									A17	Compatibility Problem		
4607	2017	TBD			CT257	Implantable cardiac pacemakers	F21	Unexpected Deterioration	A26	Insufficient Device Problem Information		

4608	2018	All other reportable incidents	47263	Cardiac resynchronization therapy implantable pacemaker	CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A07	Electrical /Electronic Property Problem		
									A0709	Device Sensing Problem		
					CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A07	Electrical /Electronic Property Problem		
									A0709	Device Sensing Problem		
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A07	Electrical /Electronic Property Problem		
									A0709	Device Sensing Problem		
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A07	Electrical /Electronic Property Problem		
									A0709	Device Sensing Problem		
4609	2017	Death	47270	Cardiac resynchronization therapy implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A26	Insufficient Device Problem Information		
4610	2018	All other reportable incidents			CT257	Implantable cardiac pacemakers	F1905	Device Revision or Replacement	A1304	Telemetry Discrepancy	29/03/2018	
									A26	Insufficient Device Problem Information	29/03/2018	
4611	2017	All other reportable incidents	47270	Cardiac resynchronization therapy implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	F1905	Device Revision or Replacement	A07	Electrical /Electronic Property Problem	25/01/2017	
									A0708	Power Problem	25/01/2017	
4612	2017	TBD	47270	Cardiac resynchronization therapy implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A07	Electrical /Electronic Property Problem		
									A0708	Power Problem		
4613	2017	TBD	35852	Single-chamber implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A07	Electrical /Electronic Property Problem		
									A0709	Device Sensing Problem		
4614	2017	TBD	35223	Endocardial/interventricular septal pacing lead	CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A05	Mechanical Problem		
									A0510	Retraction Problem		
4615	2017	All other reportable incidents			CT1684	Cardiac pacemakers and associated devices	F1001	Absence of Treatment	A15	Activation, Positioning or Separation Problem		
									A150101	Activation Failure		
4616	2018	All other reportable incidents	47270	Cardiac resynchronization therapy implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	F1905	Device Revision or Replacement	A07	Electrical /Electronic Property Problem	09/04/2018	
									A0708	Power Problem	09/04/2018	
4617	2018	All other reportable incidents			CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A09	Output Problem		
									A0907	No Device Output		
4618	2017	TBD	47270	Cardiac resynchronization therapy implantable defibrillator	CT2866	Implantable cardiac electrical stimulators			A07	Electrical /Electronic Property Problem		
									A0708	Power Problem		
4619	2018	All other reportable incidents			CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A09	Output Problem		
									A0907	No Device Output		
4620	2017	TBD	35852	Single-chamber implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A0104	Migration or Expulsion of Device		
									A09	Output Problem		
									A090404	Output Above Specifications		
4621	2018	All other reportable incidents			CT257	Implantable cardiac pacemakers	F1905	Device Revision or Replacement	A07	Electrical /Electronic Property Problem	21/08/2018	
									A0708	Power Problem	21/08/2018	
4622	2017	TBD			CT257	Implantable cardiac pacemakers	F1905	Device Revision or Replacement	A07	Electrical /Electronic Property Problem		
									A0708	Power Problem		
4623	2017	TBD	47270	Cardiac resynchronization therapy implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	F1905	Device Revision or Replacement	A13	Communication or Transmission Problem		
									A26	Insufficient Device Problem Information		
4624	2017	All other reportable incidents	47270	Cardiac resynchronization therapy implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A1104	Computer Operating System Problem		
4625	2017	Death			CT257	Implantable cardiac pacemakers	E2401	Insufficient Information	A26	Insufficient Device Problem Information		
4626	2017	TBD	35852	Single-chamber implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A0104	Migration or Expulsion of Device		
4627	2017	TBD	35223	Endocardial/interventricular septal pacing lead	CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A05	Mechanical Problem		
									A0510	Retraction Problem		
4628	2017	All other reportable incidents	47270	Cardiac resynchronization therapy implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	F1905	Device Revision or Replacement	A07	Electrical /Electronic Property Problem	24/01/2017	
									A0708	Power Problem	24/01/2017	
4629	2018	All other reportable incidents	35223	Endocardial/interventricular septal pacing lead	CT1684	Cardiac pacemakers and associated devices	F1905	Device Revision or Replacement	A07	Electrical /Electronic Property Problem		
									A0709	Device Sensing Problem		
4630	2017	TBD			CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A07	Electrical /Electronic Property Problem		
									A0709	Device Sensing Problem		
4631	2018	Death	35223	Endocardial/interventricular septal pacing lead	CT1684	Cardiac pacemakers and associated devices	E0506	Hemorrhage/Blood Loss/Bleeding	A26	Insufficient Device Problem Information		
							F21	Unexpected Deterioration	A26	Insufficient Device Problem Information		
4632	2017	TBD	35223	Endocardial/interventricular septal pacing lead	CT1684	Cardiac pacemakers and associated devices	E1906	Unspecified Infection	A26	Insufficient Device Problem Information		
4633	2017	TBD			CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A07	Electrical /Electronic Property Problem		
									A0709	Device Sensing Problem		
4634	2017	TBD	47270	Cardiac resynchronization therapy implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A07	Electrical /Electronic Property Problem		
									A0708	Power Problem		
4635	2017	TBD			CT257	Implantable cardiac pacemakers	F1905	Device Revision or Replacement	A07	Electrical /Electronic Property Problem		
									A0709	Device Sensing Problem		
4636	2017	All other reportable incidents	47270	Cardiac resynchronization therapy implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A07	Electrical /Electronic Property Problem		
									A0709	Device Sensing Problem		
4637	2018	All other reportable incidents	35223	Endocardial/interventricular septal pacing lead	CT1684	Cardiac pacemakers and associated devices	F1905	Device Revision or Replacement	A07	Electrical /Electronic Property Problem		
									A0709	Device Sensing Problem		
4638	2017	All other reportable incidents	35223	Endocardial/interventricular septal pacing lead	CT1684	Cardiac pacemakers and associated devices	F1905	Device Revision or Replacement	A05	Mechanical Problem		
									A0510	Retraction Problem		
4639	2017	TBD	35223	Endocardial/interventricular septal pacing lead	CT1684	Cardiac pacemakers and associated devices			A26	Insufficient Device Problem Information		
4640	2018	All other reportable incidents	35223	Endocardial/interventricular septal pacing lead	CT1684	Cardiac pacemakers and associated devices	F1905	Device Revision or Replacement	A05	Mechanical Problem	17/08/2018	
									A0510	Retraction Problem	17/08/2018	
4641	2017	All other reportable incidents			CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A1104	Computer Operating System Problem		
4642	2017	TBD	47270	Cardiac resynchronization therapy implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A11	Computer Software Problem		
									A1102	Application Program Problem		
4643	2017	Death			CT1684	Cardiac pacemakers and associated devices	F21	Unexpected Deterioration	A09	Output Problem		
									A0907	No Device Output		
4644	2017	All other reportable incidents	47270	Cardiac resynchronization therapy implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	F1905	Device Revision or Replacement	A07	Electrical /Electronic Property Problem	30/01/2017	
									A0708	Power Problem	30/01/2017	
4645	2017	All other reportable incidents	47270	Cardiac resynchronization therapy implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	F1905	Device Revision or Replacement	A07	Electrical /Electronic Property Problem		
									A0708	Power Problem		
4646	2018	All other reportable incidents			CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A07	Electrical /Electronic Property Problem		
									A0709	Device Sensing Problem		
4647	2017	All other reportable incidents	47270	Cardiac resynchronization therapy implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	F1905	Device Revision or Replacement	A07	Electrical /Electronic Property Problem	31/01/2017	
									A0708	Power Problem	31/01/2017	
4648	2018	All other reportable incidents	47270	Cardiac resynchronization therapy implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A07	Electrical /Electronic Property Problem		
									A0709	Device Sensing Problem		
4649	2018	All other reportable incidents	47270	Cardiac resynchronization therapy implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E2401	Insufficient Information	A07	Electrical /Electronic Property Problem		
									A0709	Device Sensing Problem		
4650	2017	TBD	35223	Endocardial/interventricular septal pacing lead	CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A07	Electrical /Electronic Property Problem		

									A0709	Device Sensing Problem		
4651	2018	All other reportable incidents			CT257	Implantable cardiac pacemakers	F1905	Device Revision or Replacement	A07	Electrical /Electronic Property Problem	31/08/2018	
									A0708	Power Problem	31/08/2018	
4652	2017	TBD	47270	Cardiac resynchronization therapy implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A16	Protective Measures Problem		
									A1601	Device Alarm System		
4653	2017	TBD	47270	Cardiac resynchronization therapy implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A16	Protective Measures Problem		
									A1602	Fail-Safe Problem		
4654	2018	All other reportable incidents	35223	Endocardial/interventricular septal pacing lead	CT1684	Cardiac pacemakers and associated devices	F1905	Device Revision or Replacement	A07	Electrical /Electronic Property Problem		
									A0709	Device Sensing Problem		
4655	2018	All other reportable incidents	35852	Single-chamber implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A07	Electrical /Electronic Property Problem		
									A0709	Device Sensing Problem		
4656	2017	All other reportable incidents	35223	Endocardial/interventricular septal pacing lead	CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A0104	Migration or Expulsion of Device		
4657	2017	TBD	47270	Cardiac resynchronization therapy implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A07	Electrical /Electronic Property Problem		
									A0709	Device Sensing Problem		
4658	2017	All other reportable incidents	47270	Cardiac resynchronization therapy implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A07	Electrical /Electronic Property Problem		
									A0709	Device Sensing Problem		
4659	2017	TBD	35852	Single-chamber implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A1304	Telemetry Discrepancy		
									A26	Insufficient Device Problem Information		
4660	2017	TBD	47270	Cardiac resynchronization therapy implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A07	Electrical /Electronic Property Problem		
									A0708	Power Problem		
4661	2018	Unanticipated serious deterioration in state of health			CT257	Implantable cardiac pacemakers			A26	Insufficient Device Problem Information		
4662	2017	All other reportable incidents	47270	Cardiac resynchronization therapy implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	F1905	Device Revision or Replacement	A07	Electrical /Electronic Property Problem	27/01/2017	
									A0708	Power Problem	27/01/2017	
4663	2017	All other reportable incidents	47270	Cardiac resynchronization therapy implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	F1905	Device Revision or Replacement	A07	Electrical /Electronic Property Problem	31/01/2017	
									A0708	Power Problem	31/01/2017	
4664	2017	All other reportable incidents	47270	Cardiac resynchronization therapy implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	F1905	Device Revision or Replacement	A07	Electrical /Electronic Property Problem	27/01/2017	
									A0708	Power Problem	27/01/2017	
4665	2017	TBD	35223	Endocardial/interventricular septal pacing lead	CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A15	Activation, Positioning or Separation Problem		
									A150101	Activation Failure		
4666	2017	All other reportable incidents	47270	Cardiac resynchronization therapy implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	F1905	Device Revision or Replacement	A07	Electrical /Electronic Property Problem	31/01/2017	
									A0708	Power Problem	31/01/2017	
4667	2017	All other reportable incidents			CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A07	Electrical /Electronic Property Problem		
									A0709	Device Sensing Problem		
4668	2017	All other reportable incidents			CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A1104	Computer Operating System Problem		
4669	2017	TBD	35223	Endocardial/interventricular septal pacing lead	CT1684	Cardiac pacemakers and associated devices			A26	Insufficient Device Problem Information		
4670	2018	All other reportable incidents	35852	Single-chamber implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A09	Output Problem		
									A090404	Output Above Specifications		
4671	2018	All other reportable incidents			CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A07	Electrical /Electronic Property Problem		
									A0709	Device Sensing Problem		
4672	2017	TBD	47270	Cardiac resynchronization therapy implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	F1905	Device Revision or Replacement	A07	Electrical /Electronic Property Problem		
									A0708	Power Problem		
4673	2018	All other reportable incidents	47270	Cardiac resynchronization therapy implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A07	Electrical /Electronic Property Problem		
									A0709	Device Sensing Problem		
4674	2017	All other reportable incidents	47270	Cardiac resynchronization therapy implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A07	Electrical /Electronic Property Problem		
									A0709	Device Sensing Problem		
4675	2017	All other reportable incidents			CT257	Implantable cardiac pacemakers	F1905	Device Revision or Replacement	A07	Electrical /Electronic Property Problem	14/06/2017	
									A0709	Device Sensing Problem	14/06/2017	
									A09	Output Problem	14/06/2017	
									A0907	No Device Output	14/06/2017	
4676	2017	TBD	35223	Endocardial/interventricular septal pacing lead	CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A04	Material Integrity Problem		
									A0405	Degraded		
4677	2017	All other reportable incidents			CT257	Implantable cardiac pacemakers	F1905	Device Revision or Replacement	A07	Electrical /Electronic Property Problem	05/01/2017	
									A0708	Power Problem	05/01/2017	
4678	2017	All other reportable incidents			CT257	Implantable cardiac pacemakers	F1905	Device Revision or Replacement	A07	Electrical /Electronic Property Problem	02/02/2017	
									A0708	Power Problem	02/02/2017	
4679	2017	All other reportable incidents	47270	Cardiac resynchronization therapy implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	F1905	Device Revision or Replacement	A07	Electrical /Electronic Property Problem	21/01/2017	
									A0708	Power Problem	21/01/2017	
4680	2017	TBD	35223	Endocardial/interventricular septal pacing lead	CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A0104	Migration or Expulsion of Device		
4681	2017	All other reportable incidents			CT1684	Cardiac pacemakers and associated devices	E2401	Insufficient Information	A07	Electrical /Electronic Property Problem		
									A0709	Device Sensing Problem		
4682	2017	All other reportable incidents	47270	Cardiac resynchronization therapy implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	F1905	Device Revision or Replacement	A07	Electrical /Electronic Property Problem	21/01/2017	
									A0708	Power Problem	21/01/2017	
4683	2017	All other reportable incidents	47270	Cardiac resynchronization therapy implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	F1905	Device Revision or Replacement	A07	Electrical /Electronic Property Problem	03/02/2017	
									A0708	Power Problem	03/02/2017	
4684	2018	All other reportable incidents			CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A07	Electrical /Electronic Property Problem		
									A0708	Power Problem		
4685	2018	All other reportable incidents			CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A09	Output Problem		
									A0907	No Device Output		
4686	2017	TBD	47270	Cardiac resynchronization therapy implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A07	Electrical /Electronic Property Problem		
									A0708	Power Problem		
4687	2017	All other reportable incidents			CT257	Implantable cardiac pacemakers	F1905	Device Revision or Replacement	A07	Electrical /Electronic Property Problem	16/06/2017	
									A0708	Power Problem	16/06/2017	
4688	2017	All other reportable incidents	35223	Endocardial/interventricular septal pacing lead	CT1684	Cardiac pacemakers and associated devices	F1905	Device Revision or Replacement	A04	Material Integrity Problem	09/11/2017	
									A0404	Crack	09/11/2017	
									A07	Electrical /Electronic Property Problem	09/11/2017	
									A0709	Device Sensing Problem	09/11/2017	
4689	2017	All other reportable incidents			CT257	Implantable cardiac pacemakers	F1905	Device Revision or Replacement	A07	Electrical /Electronic Property Problem		
									A0709	Device Sensing Problem		
4690	2017	TBD	47270	Cardiac resynchronization therapy implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A07	Electrical /Electronic Property Problem		
									A0708	Power Problem		
4691	2018	Unanticipated serious deterioration in state of health	47263	Cardiac resynchronization therapy implantable pacemaker	CT1684	Cardiac pacemakers and associated devices	F1905	Device Revision or Replacement	A07	Electrical /Electronic Property Problem	09/09/2017	
									A0709	Device Sensing Problem	09/09/2017	
					CT1685	Cardiac pacemakers	F1905	Device Revision or Replacement	A07	Electrical /Electronic Property Problem	09/09/2017	
									A0709	Device Sensing Problem	09/09/2017	
					CT257	Implantable cardiac pacemakers	F1905	Device Revision or Replacement	A07	Electrical /Electronic Property Problem	09/09/2017	

									A0709	Device Sensing Problem	09/09/2017	
					CT2866	Implantable cardiac electrical stimulators	F1905	Device Revision or Replacement	A07	Electrical /Electronic Property Problem	09/09/2017	
									A0709	Device Sensing Problem	09/09/2017	
4692	2017	TBD	35223	Endocardial/interventricular septal pacing lead	CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A15	Activation, Positioning or Separation Problem		
									A150101	Activation Failure		
4693	2018	All other reportable incidents	35223	Endocardial/interventricular septal pacing lead	CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A05	Mechanical Problem	18/07/2018	
									A0510	Retraction Problem	18/07/2018	
									A15	Activation, Positioning or Separation Problem	18/07/2018	
									A150203	Difficult or Delayed Positioning	18/07/2018	
4694	2017	Unanticipated serious deterioration in state of health	35852	Single-chamber implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E1906	Unspecified Infection	A01	Patient Device Interaction Problem		
									A17	Compatibility Problem		
4695	2018	All other reportable incidents	47270	Cardiac resynchronization therapy implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A16	Protective Measures Problem		
									A1602	Fail-Safe Problem		
4696	2017	All other reportable incidents	47270	Cardiac resynchronization therapy implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A07	Electrical /Electronic Property Problem		
									A0708	Power Problem		
4697	2017	TBD	35852	Single-chamber implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A07	Electrical /Electronic Property Problem		
									A0709	Device Sensing Problem		
4698	2017	All other reportable incidents	35852	Single-chamber implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A07	Electrical /Electronic Property Problem		
									A0709	Device Sensing Problem		
4699	2017	All other reportable incidents	47270	Cardiac resynchronization therapy implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	F1905	Device Revision or Replacement	A24	No Apparent Device Problem	06/02/2017	
									A26	Insufficient Device Problem Information	06/02/2017	
4700	2017	All other reportable incidents	47270	Cardiac resynchronization therapy implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	F1905	Device Revision or Replacement	A07	Electrical /Electronic Property Problem	02/02/2017	
									A0708	Power Problem	02/02/2017	
4701	2017	All other reportable incidents	47270	Cardiac resynchronization therapy implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	F1905	Device Revision or Replacement	A07	Electrical /Electronic Property Problem	03/02/2017	
									A0708	Power Problem	03/02/2017	
4702	2017	All other reportable incidents	47270	Cardiac resynchronization therapy implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	F1905	Device Revision or Replacement	A07	Electrical /Electronic Property Problem	10/01/2017	
									A0708	Power Problem	10/01/2017	
4703	2018	All other reportable incidents	35852	Single-chamber implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A07	Electrical /Electronic Property Problem		
									A0709	Device Sensing Problem		
4704	2017	All other reportable incidents	47270	Cardiac resynchronization therapy implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A09	Output Problem		
									A090404	Output Above Specifications		
4705	2018	All other reportable incidents			CT257	Implantable cardiac pacemakers	F1905	Device Revision or Replacement	A07	Electrical /Electronic Property Problem	20/04/2018	
									A0709	Device Sensing Problem	20/04/2018	
4706	2018	Unanticipated serious deterioration in state of health			CT257	Implantable cardiac pacemakers	E2401	Insufficient Information	A26	Insufficient Device Problem Information		
4707	2018	All other reportable incidents	35852	Single-chamber implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A07	Electrical /Electronic Property Problem		
									A0709	Device Sensing Problem		
4708	2018	All other reportable incidents			CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A07	Electrical /Electronic Property Problem		
									A0708	Power Problem		
4709	2017	All other reportable incidents	47270	Cardiac resynchronization therapy implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	F1905	Device Revision or Replacement	A07	Electrical /Electronic Property Problem	31/01/2017	
									A0708	Power Problem	31/01/2017	
4710	2018	All other reportable incidents	47270	Cardiac resynchronization therapy implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	F1905	Device Revision or Replacement	A07	Electrical /Electronic Property Problem	18/04/2018	
									A0708	Power Problem	18/04/2018	
4711	2018	All other reportable incidents	35223	Endocardial/interventricular septal pacing lead	CT1684	Cardiac pacemakers and associated devices	F1905	Device Revision or Replacement	A07	Electrical /Electronic Property Problem		
									A0709	Device Sensing Problem		
4712	2017	All other reportable incidents	47270	Cardiac resynchronization therapy implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	F1905	Device Revision or Replacement	A24	No Apparent Device Problem	31/01/2017	
									A26	Insufficient Device Problem Information	31/01/2017	
4713	2017	All other reportable incidents	47270	Cardiac resynchronization therapy implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	F1905	Device Revision or Replacement	A09	Output Problem	31/01/2017	
									A0908	Incorrect, Inadequate or Imprecise Result or Readings	31/01/2017	
									A24	No Apparent Device Problem	31/01/2017	
4714	2017	All other reportable incidents			CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A07	Electrical /Electronic Property Problem		
									A0708	Power Problem		
4715	2017	TBD	35223	Endocardial/interventricular septal pacing lead	CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A07	Electrical /Electronic Property Problem		
									A0709	Device Sensing Problem		
4716	2018	All other reportable incidents	35223	Endocardial/interventricular septal pacing lead	CT1684	Cardiac pacemakers and associated devices	F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A0104	Migration or Expulsion of Device		
									A07	Electrical /Electronic Property Problem		
									A0709	Device Sensing Problem		
4717	2017	TBD	47270	Cardiac resynchronization therapy implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A1104	Computer Operating System Problem		
4718	2018	All other reportable incidents			CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A07	Electrical /Electronic Property Problem		
									A0708	Power Problem		
4719	2018	All other reportable incidents			CT257	Implantable cardiac pacemakers	F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A07	Electrical /Electronic Property Problem		
									A0709	Device Sensing Problem		
4720	2017	TBD	35223	Endocardial/interventricular septal pacing lead	CT1684	Cardiac pacemakers and associated devices	E2401	Insufficient Information	A26	Insufficient Device Problem Information		
4721	2017	TBD	35223	Endocardial/interventricular septal pacing lead	CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A0104	Migration or Expulsion of Device		
4722	2017	TBD	47270	Cardiac resynchronization therapy implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A16	Protective Measures Problem		
									A1601	Device Alarm System		
4723	2017	All other reportable incidents	47270	Cardiac resynchronization therapy implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	F1905	Device Revision or Replacement	A07	Electrical /Electronic Property Problem	24/01/2017	
									A0708	Power Problem	24/01/2017	
4724	2017	All other reportable incidents			CT257	Implantable cardiac pacemakers	F1905	Device Revision or Replacement	A07	Electrical /Electronic Property Problem		
									A0708	Power Problem		
4725	2017	TBD	35223	Endocardial/interventricular septal pacing lead	CT1684	Cardiac pacemakers and associated devices			A26	Insufficient Device Problem Information		
4726	2017	All other reportable incidents	47263	Cardiac resynchronization therapy implantable pacemaker	CT1684	Cardiac pacemakers and associated devices	F1905	Device Revision or Replacement	A07	Electrical /Electronic Property Problem	31/10/2017	
									A0708	Power Problem	31/10/2017	
					CT1685	Cardiac pacemakers	F1905	Device Revision or Replacement	A07	Electrical /Electronic Property Problem	31/10/2017	
									A0708	Power Problem	31/10/2017	
					CT257	Implantable cardiac pacemakers	F1905	Device Revision or Replacement	A07	Electrical /Electronic Property Problem	31/10/2017	
									A0708	Power Problem	31/10/2017	
					CT2866	Implantable cardiac electrical stimulators	F1905	Device Revision or Replacement	A07	Electrical /Electronic Property Problem	31/10/2017	
									A0708	Power Problem	31/10/2017	
4727	2018	All other reportable incidents	47270	Cardiac resynchronization therapy implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	F1905	Device Revision or Replacement	A07	Electrical /Electronic Property Problem	22/08/2018	
									A0708	Power Problem	22/08/2018	
4728	2017	Death	47270	Cardiac resynchronization therapy implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	F21	Unexpected Deterioration	A07	Electrical /Electronic Property Problem		
									A0709	Device Sensing Problem		
4729	2017	TBD	47270	Cardiac resynchronization therapy implantable	CT2866	Implantable cardiac electrical stimulators	F1905	Device Revision or Replacement	A07	Electrical /Electronic Property Problem		

				defibrillator					A0708	Power Problem		
4730	2017	TBD			CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A21	Labeling, Instructions for Use or Training Problem		
									A2101	Device Markings/Labeling Problem		
4731	2017	TBD	35223	Endocardial/interventricular septal pacing lead	CT1684	Cardiac pacemakers and associated devices	F1905	Device Revision or Replacement	A07	Electrical /Electronic Property Problem		
									A0709	Device Sensing Problem		
4732	2018	All other reportable incidents	35223	Endocardial/interventricular septal pacing lead	CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A05	Mechanical Problem		
									A0510	Retraction Problem		
4733	2017	All other reportable incidents			CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A07	Electrical /Electronic Property Problem	09/11/2017	
									A0708	Power Problem	09/11/2017	
4734	2017	All other reportable incidents	47270	Cardiac resynchronization therapy implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	F1905	Device Revision or Replacement	A07	Electrical /Electronic Property Problem		
									A0708	Power Problem		
4735	2017	All other reportable incidents	47270	Cardiac resynchronization therapy implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A13	Communication or Transmission Problem		
									A26	Insufficient Device Problem Information		
4736	2017	All other reportable incidents			CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A07	Electrical /Electronic Property Problem		
									A0708	Power Problem		
4737	2018	All other reportable incidents	47270	Cardiac resynchronization therapy implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	F1905	Device Revision or Replacement	A04	Material Integrity Problem	06/04/2018	
									A0404	Crack	06/04/2018	
4738	2018	All other reportable incidents			CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A07	Electrical /Electronic Property Problem		
									A0709	Device Sensing Problem		
									A11	Computer Software Problem		
									A1102	Application Program Problem		
4739	2018	All other reportable incidents	47270	Cardiac resynchronization therapy implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	F1905	Device Revision or Replacement	A07	Electrical /Electronic Property Problem	04/07/2018	
									A0709	Device Sensing Problem	04/07/2018	
4740	2017	All other reportable incidents	35223	Endocardial/interventricular septal pacing lead	CT1684	Cardiac pacemakers and associated devices	E2401	Insufficient Information	A26	Insufficient Device Problem Information		
4741	2017	All other reportable incidents			CT1684	Cardiac pacemakers and associated devices	F1905	Device Revision or Replacement	A07	Electrical /Electronic Property Problem		
									A0708	Power Problem		
4742	2018	All other reportable incidents	47270	Cardiac resynchronization therapy implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A07	Electrical /Electronic Property Problem		
									A0708	Power Problem		
4743	2018	All other reportable incidents			CT257	Implantable cardiac pacemakers	F1905	Device Revision or Replacement	A07	Electrical /Electronic Property Problem	10/09/2018	
									A0709	Device Sensing Problem	10/09/2018	
4744	2018	All other reportable incidents			CT257	Implantable cardiac pacemakers	F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A26	Insufficient Device Problem Information		
4745	2018	All other reportable incidents			CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A07	Electrical /Electronic Property Problem		
									A0708	Power Problem		
4746	2017	All other reportable incidents			CT257	Implantable cardiac pacemakers	F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A07	Electrical /Electronic Property Problem		
									A0709	Device Sensing Problem		
4747	2017	TBD	35223	Endocardial/interventricular septal pacing lead	CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A15	Activation, Positioning or Separation Problem		
									A150101	Activation Failure		
4748	2017	All other reportable incidents	35223	Endocardial/interventricular septal pacing lead	CT1684	Cardiac pacemakers and associated devices	E2401	Insufficient Information	A26	Insufficient Device Problem Information		
4749	2017	TBD	35852	Single-chamber implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A07	Electrical /Electronic Property Problem		
									A0709	Device Sensing Problem		
4750	2017	All other reportable incidents	35852	Single-chamber implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A07	Electrical /Electronic Property Problem		
									A0709	Device Sensing Problem		
4751	2017	All other reportable incidents	35223	Endocardial/interventricular septal pacing lead	CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A05	Mechanical Problem		
									A0510	Retraction Problem		
4752	2018	All other reportable incidents	47270	Cardiac resynchronization therapy implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A07	Electrical /Electronic Property Problem		
									A0709	Device Sensing Problem		
4753	2018	All other reportable incidents	47270	Cardiac resynchronization therapy implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A07	Electrical /Electronic Property Problem		
									A0709	Device Sensing Problem		
4754	2017	TBD			CT257	Implantable cardiac pacemakers	F1905	Device Revision or Replacement	A07	Electrical /Electronic Property Problem		
									A0709	Device Sensing Problem		
									A1304	Telemetry Discrepancy		
									A26	Insufficient Device Problem Information		
4755	2018	All other reportable incidents	35223	Endocardial/interventricular septal pacing lead	CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A05	Mechanical Problem		
									A0510	Retraction Problem		
4756	2017	All other reportable incidents	47270	Cardiac resynchronization therapy implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	F1905	Device Revision or Replacement	A07	Electrical /Electronic Property Problem	20/06/2017	
									A0708	Power Problem	20/06/2017	
4757	2018	All other reportable incidents	47270	Cardiac resynchronization therapy implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A07	Electrical /Electronic Property Problem		
									A0709	Device Sensing Problem		
4758	2017	All other reportable incidents	60789	Intracardiac pacemaker	CT130	Catheters and associated devices	F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A09	Output Problem		
									A0907	No Device Output		
							F1905	Device Revision or Replacement	A09	Output Problem		
									A0907	No Device Output		
					CT145	Cardiovascular devices	F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A09	Output Problem		
									A0907	No Device Output		
							F1905	Device Revision or Replacement	A09	Output Problem		
									A0907	No Device Output		
					CT1684	Cardiac pacemakers and associated devices	F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A09	Output Problem		
									A0907	No Device Output		
							F1905	Device Revision or Replacement	A09	Output Problem		
									A0907	No Device Output		
					CT1685	Cardiac pacemakers	F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A09	Output Problem		
									A0907	No Device Output		
							F1905	Device Revision or Replacement	A09	Output Problem		
									A0907	No Device Output		
					CT233	Surgical	F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A09	Output Problem		
									A0907	No Device Output		
							F1905	Device Revision or Replacement	A09	Output Problem		
									A0907	No Device Output		
					CT2406	Implantable	F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A09	Output Problem		
									A0907	No Device Output		
							F1905	Device Revision or Replacement	A09	Output Problem		
									A0907	No Device Output		
					CT2407	Active implantable	F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A09	Output Problem		
									A0907	No Device Output		
							F1905	Device Revision or Replacement	A09	Output Problem		

									A0907	No Device Output		
					CT257	Implantable cardiac pacemakers	F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A09	Output Problem		
									A0907	No Device Output		
							F1905	Device Revision or Replacement	A09	Output Problem		
									A0907	No Device Output		
					CT2727	Active implant devices/ancillaries	F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A09	Output Problem		
									A0907	No Device Output		
							F1905	Device Revision or Replacement	A09	Output Problem		
									A0907	No Device Output		
					CT2750	Implantable cardiac electrical stimulators and associated devices	F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A09	Output Problem		
									A0907	No Device Output		
							F1905	Device Revision or Replacement	A09	Output Problem		
									A0907	No Device Output		
					CT2818	Catheterization/Cannulation/Fluid path devices	F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A09	Output Problem		
									A0907	No Device Output		
							F1905	Device Revision or Replacement	A09	Output Problem		
									A0907	No Device Output		
					CT2866	Implantable cardiac electrical stimulators	F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A09	Output Problem		
									A0907	No Device Output		
							F1905	Device Revision or Replacement	A09	Output Problem		
									A0907	No Device Output		
					CT2985	Active	F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A09	Output Problem		
									A0907	No Device Output		
							F1905	Device Revision or Replacement	A09	Output Problem		
									A0907	No Device Output		
					CT3001	Implantable metal	F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A09	Output Problem		
									A0907	No Device Output		
							F1905	Device Revision or Replacement	A09	Output Problem		
									A0907	No Device Output		
					CT3002	Implantable synthetic polymer	F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A09	Output Problem		
									A0907	No Device Output		
							F1905	Device Revision or Replacement	A09	Output Problem		
									A0907	No Device Output		
					CT3006	Implantable inorganic material	F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A09	Output Problem		
									A0907	No Device Output		
							F1905	Device Revision or Replacement	A09	Output Problem		
									A0907	No Device Output		
					CT314	Electrically powered	F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A09	Output Problem		
									A0907	No Device Output		
							F1905	Device Revision or Replacement	A09	Output Problem		
									A0907	No Device Output		
					CT321	Long-term surgical invasive	F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A09	Output Problem		
									A0907	No Device Output		
							F1905	Device Revision or Replacement	A09	Output Problem		
									A0907	No Device Output		
					CT981	Single use	F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A09	Output Problem		
									A0907	No Device Output		
							F1905	Device Revision or Replacement	A09	Output Problem		
									A0907	No Device Output		
					CT983	Surgical invasive	F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A09	Output Problem		
									A0907	No Device Output		
							F1905	Device Revision or Replacement	A09	Output Problem		
									A0907	No Device Output		
4759	2018	All other reportable incidents	47270	Cardiac resynchronization therapy implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A07	Electrical /Electronic Property Problem		
									A0709	Device Sensing Problem		
									A070909	Oversensing		
4760	2017	TBD			CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A07	Electrical /Electronic Property Problem		
									A0708	Power Problem		
4761	2017	TBD	35852	Single-chamber implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A11	Computer Software Problem		
									A1102	Application Program Problem		
									A1304	Telemetry Discrepancy		
									A26	Insufficient Device Problem Information		
4762	2017	TBD	35852	Single-chamber implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E1906	Unspecified Infection	A01	Patient Device Interaction Problem		
									A17	Compatibility Problem		
							F1905	Device Revision or Replacement	A01	Patient Device Interaction Problem		
									A17	Compatibility Problem		
4763	2017	All other reportable incidents	35223	Endocardial/interventricular septal pacing lead	CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A15	Activation, Positioning or Separation Problem		
									A150203	Difficult or Delayed Positioning		
4764	2018	All other reportable incidents			CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A09	Output Problem		
									A0907	No Device Output		
4765	2018	All other reportable incidents			CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A05	Mechanical Problem		
									A0504	Leak/Splash		
									A11	Computer Software Problem		
									A1102	Application Program Problem		
4766	2017	All other reportable incidents			CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A26	Insufficient Device Problem Information		
4767	2018	All other reportable incidents	47270	Cardiac resynchronization therapy implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A07	Electrical /Electronic Property Problem		
									A0708	Power Problem		
4768	2017	TBD	47270	Cardiac resynchronization therapy implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	F1905	Device Revision or Replacement	A07	Electrical /Electronic Property Problem		
									A0708	Power Problem		
4769	2017	TBD	35223	Endocardial/interventricular septal pacing lead	CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A15	Activation, Positioning or Separation Problem		
									A150203	Difficult or Delayed Positioning		
4770	2017	All other reportable incidents			CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A07	Electrical /Electronic Property Problem		
									A0708	Power Problem		
4771	2017	All other reportable incidents	35852	Single-chamber implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A07	Electrical /Electronic Property Problem		
									A0709	Device Sensing Problem		
4772	2017	TBD	35852	Single-chamber implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A07	Electrical /Electronic Property Problem		
									A0709	Device Sensing Problem		
4773	2018	All other reportable incidents	35223	Endocardial/interventricular septal pacing lead	CT1684	Cardiac pacemakers and associated devices	F14	Prolonged Episode of Care	A05	Mechanical Problem	18/07/2018	

									A0510	Retraction Problem	18/07/2018	
4774	2017	All other reportable incidents			CT257	Implantable cardiac pacemakers	F1905	Device Revision or Replacement	A1304	Telemetry Discrepancy	13/01/2017	
									A26	Insufficient Device Problem Information	13/01/2017	
4775	2017	All other reportable incidents			CT257	Implantable cardiac pacemakers	E2401	Insufficient Information	A26	Insufficient Device Problem Information		
4776	2017	TBD	35223	Endocardial/interventricular septal pacing lead	CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A0104	Migration or Expulsion of Device		
4777	2018	Unanticipated serious deterioration in state of health			CT257	Implantable cardiac pacemakers			A26	Insufficient Device Problem Information		
4778	2017	TBD	35223	Endocardial/interventricular septal pacing lead	CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A15	Activation, Positioning or Separation Problem		
									A150203	Difficult or Delayed Positioning		
4779	2017	TBD	35223	Endocardial/interventricular septal pacing lead	CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A05	Mechanical Problem		
									A0512	Unintended Movement		
4780	2017	TBD	35852	Single-chamber implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A07	Electrical /Electronic Property Problem		
									A0709	Device Sensing Problem		
4781	2017	All other reportable incidents	35223	Endocardial/interventricular septal pacing lead	CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A05	Mechanical Problem		
									A0510	Retraction Problem		
4782	2017	All other reportable incidents	47270	Cardiac resynchronization therapy implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E1906	Unspecified Infection	A01	Patient Device Interaction Problem	27/11/2017	
									A17	Compatibility Problem	27/11/2017	
							F1905	Device Revision or Replacement	A01	Patient Device Interaction Problem	27/11/2017	
									A17	Compatibility Problem	27/11/2017	
4783	2017	TBD			CT257	Implantable cardiac pacemakers			A26	Insufficient Device Problem Information		
4784	2017	TBD	35223	Endocardial/interventricular septal pacing lead	CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A07	Electrical /Electronic Property Problem		
									A0709	Device Sensing Problem		
4785	2017	All other reportable incidents			CT257	Implantable cardiac pacemakers	F1905	Device Revision or Replacement	A13	Communication or Transmission Problem	04/12/2017	
									A26	Insufficient Device Problem Information	04/12/2017	
4786	2017	All other reportable incidents			CT257	Implantable cardiac pacemakers	F1905	Device Revision or Replacement	A07	Electrical /Electronic Property Problem	02/03/2017	
									A0709	Device Sensing Problem	02/03/2017	
4787	2017	All other reportable incidents			CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A0902	Display or Visual Feedback Problem		
									A1104	Computer Operating System Problem		
									A26	Insufficient Device Problem Information		
4788	2017	TBD			CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A26	Insufficient Device Problem Information		
4789	2017	TBD			CT257	Implantable cardiac pacemakers	F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A26	Insufficient Device Problem Information		
4790	2017	All other reportable incidents			CT257	Implantable cardiac pacemakers	F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A11	Computer Software Problem		
									A1102	Application Program Problem		
4791	2018	All other reportable incidents	35223	Endocardial/interventricular septal pacing lead	CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A05	Mechanical Problem	12/09/2018	
									A0510	Retraction Problem	12/09/2018	
4792	2018	All other reportable incidents			CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A09	Output Problem		
									A0907	No Device Output		
4793	2018	All other reportable incidents			CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A07	Electrical /Electronic Property Problem		
									A0709	Device Sensing Problem		
4794	2017	Death			CT257	Implantable cardiac pacemakers	F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A26	Insufficient Device Problem Information		
							F21	Unexpected Deterioration	A26	Insufficient Device Problem Information		
4795	2018	All other reportable incidents	47270	Cardiac resynchronization therapy implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A07	Electrical /Electronic Property Problem		
									A0708	Power Problem		
4796	2018	All other reportable incidents	35852	Single-chamber implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A09	Output Problem		
									A090404	Output Above Specifications		
									A0908	Incorrect, Inadequate or Imprecise Result or Readings		
4797	2017	TBD			CT257	Implantable cardiac pacemakers	F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A26	Insufficient Device Problem Information		
4798	2017	All other reportable incidents	47270	Cardiac resynchronization therapy implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A07	Electrical /Electronic Property Problem		
									A0708	Power Problem		
4799	2018	All other reportable incidents	35852	Single-chamber implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A07	Electrical /Electronic Property Problem		
									A0709	Device Sensing Problem		
4800	2018	All other reportable incidents	35852	Single-chamber implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	F1001	Absence of Treatment	A12	Connection Problem		
									A1206	Misconnection		
4801	2018	All other reportable incidents	35852	Single-chamber implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E2104	Electric Shock	A09	Output Problem		
									A090404	Output Above Specifications		
4802	2018	All other reportable incidents			CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A07	Electrical /Electronic Property Problem		
									A0709	Device Sensing Problem		
4803	2017	All other reportable incidents			CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A07	Electrical /Electronic Property Problem	01/03/2017	
									A0708	Power Problem	01/03/2017	
4804	2018	All other reportable incidents			CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A09	Output Problem		
									A0907	No Device Output		
4805	2018	All other reportable incidents			CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A07	Electrical /Electronic Property Problem		
									A0708	Power Problem		
4806	2017	TBD	47270	Cardiac resynchronization therapy implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A07	Electrical /Electronic Property Problem		
									A0709	Device Sensing Problem		
4807	2017	All other reportable incidents	47270	Cardiac resynchronization therapy implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	F1905	Device Revision or Replacement	A07	Electrical /Electronic Property Problem	16/02/2017	
									A0708	Power Problem	16/02/2017	
4808	2017	All other reportable incidents	35852	Single-chamber implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E2104	Electric Shock	A09	Output Problem		
									A090404	Output Above Specifications		
4809	2017	Death	47270	Cardiac resynchronization therapy implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	F21	Unexpected Deterioration	A26	Insufficient Device Problem Information		
4810	2017	All other reportable incidents	47270	Cardiac resynchronization therapy implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	F1905	Device Revision or Replacement	A07	Electrical /Electronic Property Problem	26/06/2017	
									A0708	Power Problem	26/06/2017	
4811	2017	TBD			CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A07	Electrical /Electronic Property Problem		
									A0708	Power Problem		
4812	2018	All other reportable incidents	47270	Cardiac resynchronization therapy implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A26	Insufficient Device Problem Information		
4813	2018	All other reportable incidents	35852	Single-chamber implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A07	Electrical /Electronic Property Problem		
									A0709	Device Sensing Problem		
4814	2018	All other reportable incidents	47270	Cardiac resynchronization therapy implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A07	Electrical /Electronic Property Problem		
									A0709	Device Sensing Problem		
4815	2018	All other reportable incidents	47270	Cardiac resynchronization therapy implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A07	Electrical /Electronic Property Problem		
									A0709	Device Sensing Problem		
4816	2018	All other reportable incidents	47270	Cardiac resynchronization therapy implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A09	Output Problem		
									A090404	Output Above Specifications		
4817	2018	All other reportable incidents	35223	Endocardial/interventricular septal pacing lead	CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A07	Electrical /Electronic Property Problem		
									A0709	Device Sensing Problem		

4818	2017	TBD	35223	Endocardial/interventricular septal pacing lead	CT1684	Cardiac pacemakers and associated devices	F10	Inadequate/inappropriate Treatment or Diagnostic Exposure	A07	Electrical /Electronic Property Problem		
									A0709	Device Sensing Problem		
4819	2017	All other reportable incidents			CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A07	Electrical /Electronic Property Problem		
									A0709	Device Sensing Problem		
4820	2018	All other reportable incidents	60789	Intracardiac pacemaker	CT130	Catheters and associated devices	E1906	Unspecified Infection	A0104	Migration or Expulsion of Device		
							E2114	Perforation	A0104	Migration or Expulsion of Device		
							F10	Inadequate/inappropriate Treatment or Diagnostic Exposure	A0104	Migration or Expulsion of Device		
					CT145	Cardiovascular devices	E1906	Unspecified Infection	A0104	Migration or Expulsion of Device		
							E2114	Perforation	A0104	Migration or Expulsion of Device		
							F10	Inadequate/inappropriate Treatment or Diagnostic Exposure	A0104	Migration or Expulsion of Device		
					CT1684	Cardiac pacemakers and associated devices	E1906	Unspecified Infection	A0104	Migration or Expulsion of Device		
							E2114	Perforation	A0104	Migration or Expulsion of Device		
							F10	Inadequate/inappropriate Treatment or Diagnostic Exposure	A0104	Migration or Expulsion of Device		
					CT1685	Cardiac pacemakers	E1906	Unspecified Infection	A0104	Migration or Expulsion of Device		
							E2114	Perforation	A0104	Migration or Expulsion of Device		
							F10	Inadequate/inappropriate Treatment or Diagnostic Exposure	A0104	Migration or Expulsion of Device		
					CT233	Surgical	E1906	Unspecified Infection	A0104	Migration or Expulsion of Device		
							E2114	Perforation	A0104	Migration or Expulsion of Device		
							F10	Inadequate/inappropriate Treatment or Diagnostic Exposure	A0104	Migration or Expulsion of Device		
					CT2406	Implantable	E1906	Unspecified Infection	A0104	Migration or Expulsion of Device		
							E2114	Perforation	A0104	Migration or Expulsion of Device		
							F10	Inadequate/inappropriate Treatment or Diagnostic Exposure	A0104	Migration or Expulsion of Device		
					CT2407	Active implantable	E1906	Unspecified Infection	A0104	Migration or Expulsion of Device		
							E2114	Perforation	A0104	Migration or Expulsion of Device		
							F10	Inadequate/inappropriate Treatment or Diagnostic Exposure	A0104	Migration or Expulsion of Device		
					CT257	Implantable cardiac pacemakers	E1906	Unspecified Infection	A0104	Migration or Expulsion of Device		
							E2114	Perforation	A0104	Migration or Expulsion of Device		
							F10	Inadequate/inappropriate Treatment or Diagnostic Exposure	A0104	Migration or Expulsion of Device		
					CT2727	Active implant devices/ancillaries	E1906	Unspecified Infection	A0104	Migration or Expulsion of Device		
							E2114	Perforation	A0104	Migration or Expulsion of Device		
							F10	Inadequate/inappropriate Treatment or Diagnostic Exposure	A0104	Migration or Expulsion of Device		
					CT2750	Implantable cardiac electrical stimulators and associated devices	E1906	Unspecified Infection	A0104	Migration or Expulsion of Device		
							E2114	Perforation	A0104	Migration or Expulsion of Device		
							F10	Inadequate/inappropriate Treatment or Diagnostic Exposure	A0104	Migration or Expulsion of Device		
					CT2818	Catheterization/Cannulation/Fluid path devices	E1906	Unspecified Infection	A0104	Migration or Expulsion of Device		
							E2114	Perforation	A0104	Migration or Expulsion of Device		
							F10	Inadequate/inappropriate Treatment or Diagnostic Exposure	A0104	Migration or Expulsion of Device		
					CT2866	Implantable cardiac electrical stimulators	E1906	Unspecified Infection	A0104	Migration or Expulsion of Device		
							E2114	Perforation	A0104	Migration or Expulsion of Device		
							F10	Inadequate/inappropriate Treatment or Diagnostic Exposure	A0104	Migration or Expulsion of Device		
					CT2985	Active	E1906	Unspecified Infection	A0104	Migration or Expulsion of Device		
							E2114	Perforation	A0104	Migration or Expulsion of Device		
							F10	Inadequate/inappropriate Treatment or Diagnostic Exposure	A0104	Migration or Expulsion of Device		
					CT3001	Implantable metal	E1906	Unspecified Infection	A0104	Migration or Expulsion of Device		
							E2114	Perforation	A0104	Migration or Expulsion of Device		
							F10	Inadequate/inappropriate Treatment or Diagnostic Exposure	A0104	Migration or Expulsion of Device		
					CT3002	Implantable synthetic polymer	E1906	Unspecified Infection	A0104	Migration or Expulsion of Device		
							E2114	Perforation	A0104	Migration or Expulsion of Device		
							F10	Inadequate/inappropriate Treatment or Diagnostic Exposure	A0104	Migration or Expulsion of Device		
					CT3006	Implantable inorganic material	E1906	Unspecified Infection	A0104	Migration or Expulsion of Device		
							E2114	Perforation	A0104	Migration or Expulsion of Device		
							F10	Inadequate/inappropriate Treatment or Diagnostic Exposure	A0104	Migration or Expulsion of Device		
					CT314	Electrically powered	E1906	Unspecified Infection	A0104	Migration or Expulsion of Device		
							E2114	Perforation	A0104	Migration or Expulsion of Device		
							F10	Inadequate/inappropriate Treatment or Diagnostic Exposure	A0104	Migration or Expulsion of Device		
					CT321	Long-term surgical invasive	E1906	Unspecified Infection	A0104	Migration or Expulsion of Device		
							E2114	Perforation	A0104	Migration or Expulsion of Device		
							F10	Inadequate/inappropriate Treatment or Diagnostic Exposure	A0104	Migration or Expulsion of Device		
					CT981	Single use	E1906	Unspecified Infection	A0104	Migration or Expulsion of Device		
							E2114	Perforation	A0104	Migration or Expulsion of Device		
							F10	Inadequate/inappropriate Treatment or Diagnostic Exposure	A0104	Migration or Expulsion of Device		
					CT983	Surgical invasive	E1906	Unspecified Infection	A0104	Migration or Expulsion of Device		
							E2114	Perforation	A0104	Migration or Expulsion of Device		
							F10	Inadequate/inappropriate Treatment or Diagnostic Exposure	A0104	Migration or Expulsion of Device		
4821	2017	All other reportable incidents	35223	Endocardial/interventricular septal pacing lead	CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A15	Activation, Positioning or Separation Problem	24/11/2017	
									A150101	Activation Failure	24/11/2017	
4822	2017	TBD			CT257	Implantable cardiac pacemakers	F1905	Device Revision or Replacement	A07	Electrical /Electronic Property Problem		
									A0708	Power Problem		
4823	2018	All other reportable incidents	47270	Cardiac resynchronization therapy implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	F1905	Device Revision or Replacement	A07	Electrical /Electronic Property Problem	23/04/2018	
									A0708	Power Problem	23/04/2018	
4824	2017	All other reportable incidents			CT257	Implantable cardiac pacemakers	F1905	Device Revision or Replacement	A07	Electrical /Electronic Property Problem		
									A0709	Device Sensing Problem		
4825	2017	All other reportable incidents	35223	Endocardial/interventricular septal pacing lead	CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A15	Activation, Positioning or Separation Problem		
									A150101	Activation Failure		
4826	2017	All other reportable incidents	47270	Cardiac resynchronization therapy implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	F1905	Device Revision or Replacement	A07	Electrical /Electronic Property Problem	12/08/2016	
									A0708	Power Problem	12/08/2016	
4827	2017	TBD	35223	Endocardial/interventricular septal pacing lead	CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A15	Activation, Positioning or Separation Problem		
									A150101	Activation Failure		
4828	2017	All other reportable incidents	35223	Endocardial/interventricular septal pacing lead	CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A07	Electrical /Electronic Property Problem		
									A0709	Device Sensing Problem		
4829	2017	All other reportable incidents	35223	Endocardial/interventricular septal pacing lead	CT1684	Cardiac pacemakers and associated devices	E2401	Insufficient Information	A26	Insufficient Device Problem Information		
4830	2018	Unanticipated serious deterioration in state of health	35223	Endocardial/interventricular septal pacing lead	CT1684	Cardiac pacemakers and associated devices	F1905	Device Revision or Replacement	A07	Electrical /Electronic Property Problem		
									A0709	Device Sensing Problem		
4831	2017	All other reportable incidents	35852	Single-chamber implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E1906	Unspecified Infection	A01	Patient Device Interaction Problem		
									A17	Compatibility Problem		
							F1905	Device Revision or Replacement	A01	Patient Device Interaction Problem		
									A17	Compatibility Problem		

4832	2017	All other reportable incidents	35223	Endocardial/interventricular septal pacing lead	CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A09	Output Problem		
									A0908	Incorrect, Inadequate or Imprecise Result or Readings		
4833	2018	All other reportable incidents			CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A07	Electrical /Electronic Property Problem		
									A0708	Power Problem		
4834	2017	TBD	47270	Cardiac resynchronization therapy implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A07	Electrical /Electronic Property Problem		
									A0708	Power Problem		
4835	2017	TBD	35223	Endocardial/interventricular septal pacing lead	CT1684	Cardiac pacemakers and associated devices	E0402	Hypersensitivity/Allergic Reaction	A26	Insufficient Device Problem Information		
4836	2018	All other reportable incidents			CT257	Implantable cardiac pacemakers	F1905	Device Revision or Replacement	A07	Electrical /Electronic Property Problem	03/09/2018	
									A0708	Power Problem	03/09/2018	
4837	2017	TBD	35223	Endocardial/interventricular septal pacing lead	CT1684	Cardiac pacemakers and associated devices	F14	Prolonged Episode of Care	A04	Material Integrity Problem		
									A0404	Crack		
4838	2017	All other reportable incidents	47270	Cardiac resynchronization therapy implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A07	Electrical /Electronic Property Problem	07/12/2017	
									A0708	Power Problem	07/12/2017	
4839	2018	All other reportable incidents	35852	Single-chamber implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A09	Output Problem		
									A090404	Output Above Specifications		
4840	2018	All other reportable incidents	47270	Cardiac resynchronization therapy implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A07	Electrical /Electronic Property Problem		
									A0709	Device Sensing Problem		
4841	2018	All other reportable incidents	47270	Cardiac resynchronization therapy implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A07	Electrical /Electronic Property Problem		
									A0709	Device Sensing Problem		
4842	2017	All other reportable incidents	47270	Cardiac resynchronization therapy implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	F1905	Device Revision or Replacement	A07	Electrical /Electronic Property Problem		
									A0708	Power Problem		
4843	2017	All other reportable incidents	35852	Single-chamber implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A11	Computer Software Problem		
									A1102	Application Program Problem		
4844	2017	All other reportable incidents	35223	Endocardial/interventricular septal pacing lead	CT1684	Cardiac pacemakers and associated devices	E2401	Insufficient Information	A26	Insufficient Device Problem Information		
4845	2018	All other reportable incidents	47263	Cardiac resynchronization therapy implantable pacemaker	CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A05	Mechanical Problem	01/08/2012	
									A0506	Mechanical Jam	01/08/2012	
					CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A05	Mechanical Problem	01/08/2012	
									A0506	Mechanical Jam	01/08/2012	
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A05	Mechanical Problem	01/08/2012	
									A0506	Mechanical Jam	01/08/2012	
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A05	Mechanical Problem	01/08/2012	
									A0506	Mechanical Jam	01/08/2012	
4846	2017	TBD			CT257	Implantable cardiac pacemakers	F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A07	Electrical /Electronic Property Problem		
									A0708	Power Problem		
4847	2017	All other reportable incidents	47263	Cardiac resynchronization therapy implantable pacemaker	CT1684	Cardiac pacemakers and associated devices	F21	Unexpected Deterioration	A09	Output Problem	10/01/2017	
									A0907	No Device Output	10/01/2017	
					CT1685	Cardiac pacemakers	F21	Unexpected Deterioration	A09	Output Problem	10/01/2017	
									A0907	No Device Output	10/01/2017	
					CT257	Implantable cardiac pacemakers	F21	Unexpected Deterioration	A09	Output Problem	10/01/2017	
									A0907	No Device Output	10/01/2017	
					CT2866	Implantable cardiac electrical stimulators	F21	Unexpected Deterioration	A09	Output Problem	10/01/2017	
									A0907	No Device Output	10/01/2017	
4848	2018	All other reportable incidents	35223	Endocardial/interventricular septal pacing lead	CT1684	Cardiac pacemakers and associated devices	E2114	Perforation	A0104	Migration or Expulsion of Device	09/04/2018	
							F1905	Device Revision or Replacement	A0104	Migration or Expulsion of Device	09/04/2018	
4849	2017	TBD	35223	Endocardial/interventricular septal pacing lead	CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A15	Activation, Positioning or Separation Problem		
									A150101	Activation Failure		
4850	2017	All other reportable incidents			CT257	Implantable cardiac pacemakers	F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A07	Electrical /Electronic Property Problem		
									A0709	Device Sensing Problem		
4851	2017	All other reportable incidents			CT1684	Cardiac pacemakers and associated devices	E2401	Insufficient Information	A09	Output Problem		
									A0908	Incorrect, Inadequate or Imprecise Result or Readings		
4852	2017	Death			CT257	Implantable cardiac pacemakers	F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A09	Output Problem		
									A090404	Output Above Specifications		
4853	2017	All other reportable incidents	47270	Cardiac resynchronization therapy implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A07	Electrical /Electronic Property Problem	16/11/2017	
									A0708	Power Problem	16/11/2017	
4854	2017	All other reportable incidents	47270	Cardiac resynchronization therapy implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	F1905	Device Revision or Replacement	A07	Electrical /Electronic Property Problem	08/12/2017	
									A0708	Power Problem	08/12/2017	
4855	2018	All other reportable incidents	47270	Cardiac resynchronization therapy implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A07	Electrical /Electronic Property Problem		
									A0709	Device Sensing Problem		
4856	2017	All other reportable incidents			CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A07	Electrical /Electronic Property Problem	01/02/2017	
									A0708	Power Problem	01/02/2017	
4857	2017	TBD	47270	Cardiac resynchronization therapy implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A07	Electrical /Electronic Property Problem		
									A0708	Power Problem		
4858	2018	All other reportable incidents			CT257	Implantable cardiac pacemakers	F1905	Device Revision or Replacement	A07	Electrical /Electronic Property Problem	27/04/2018	
									A0708	Power Problem	27/04/2018	
									A13	Communication or Transmission Problem	27/04/2018	
									A26	Insufficient Device Problem Information	27/04/2018	
4859	2018	All other reportable incidents	47270	Cardiac resynchronization therapy implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A11	Computer Software Problem		
									A1102	Application Program Problem		
4860	2017	TBD			CT257	Implantable cardiac pacemakers	F21	Unexpected Deterioration	A09	Output Problem		
									A0907	No Device Output		
4861	2017	TBD			CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A07	Electrical /Electronic Property Problem		
									A0708	Power Problem		
4862	2017	All other reportable incidents			CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A07	Electrical /Electronic Property Problem		
									A0708	Power Problem		
4863	2017	All other reportable incidents	35223	Endocardial/interventricular septal pacing lead	CT1684	Cardiac pacemakers and associated devices	F1905	Device Revision or Replacement	A07	Electrical /Electronic Property Problem	11/07/2017	
									A0709	Device Sensing Problem	11/07/2017	
4864	2017	All other reportable incidents	35852	Single-chamber implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A07	Electrical /Electronic Property Problem		
									A0709	Device Sensing Problem		
4865	2017	TBD	47270	Cardiac resynchronization therapy implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	F1905	Device Revision or Replacement	A07	Electrical /Electronic Property Problem		
									A0708	Power Problem		
4866	2017	TBD	47270	Cardiac resynchronization therapy implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A07	Electrical /Electronic Property Problem		
									A0708	Power Problem		
4867	2018	All other reportable incidents			CT257	Implantable cardiac pacemakers	E2007	Fall	A26	Insufficient Device Problem Information		
4868	2017	TBD			CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A16	Protective Measures Problem		
									A1606	Premature Indicator Activation		

									A160601	Premature Elective Replacement Indicator		
4869	2018	Unanticipated serious deterioration in state of health	47270	Cardiac resynchronization therapy implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	F14	Prolonged Episode of Care	A01	Patient Device Interaction Problem		
									A17	Compatibility Problem		
4870	2017	Death	47270	Cardiac resynchronization therapy implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A09	Output Problem		
									A0907	No Device Output		
4871	2017	Unanticipated serious deterioration in state of health			CT257	Implantable cardiac pacemakers	F1905	Device Revision or Replacement	A07	Electrical /Electronic Property Problem		
									A0709	Device Sensing Problem		
4872	2018	All other reportable incidents			CT1684	Cardiac pacemakers and associated devices	F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A09	Output Problem		
									A0907	No Device Output		
4873	2018	All other reportable incidents	47270	Cardiac resynchronization therapy implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A07	Electrical /Electronic Property Problem	24/04/2018	
									A0708	Power Problem	24/04/2018	
4874	2017	All other reportable incidents	47270	Cardiac resynchronization therapy implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A17	Compatibility Problem		
									A1701	Component or Accessory Incompatibility		
4875	2017	All other reportable incidents	47263	Cardiac resynchronization therapy implantable pacemaker	CT1684	Cardiac pacemakers and associated devices	F1905	Device Revision or Replacement	A09	Output Problem		
									A0908	Incorrect, Inadequate or Imprecise Result or Readings		
					CT1685	Cardiac pacemakers	F1905	Device Revision or Replacement	A09	Output Problem		
									A0908	Incorrect, Inadequate or Imprecise Result or Readings		
					CT257	Implantable cardiac pacemakers	F1905	Device Revision or Replacement	A09	Output Problem		
									A0908	Incorrect, Inadequate or Imprecise Result or Readings		
					CT2866	Implantable cardiac electrical stimulators	F1905	Device Revision or Replacement	A09	Output Problem		
									A0908	Incorrect, Inadequate or Imprecise Result or Readings		
4876	2017	TBD	35852	Single-chamber implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A07	Electrical /Electronic Property Problem		
									A0709	Device Sensing Problem		
4877	2017	All other reportable incidents	35223	Endocardial/interventricular septal pacing lead	CT1684	Cardiac pacemakers and associated devices	F1905	Device Revision or Replacement	A07	Electrical /Electronic Property Problem		
									A0709	Device Sensing Problem		
4878	2018	Unanticipated serious deterioration in state of health			CT257	Implantable cardiac pacemakers	F1001	Absence of Treatment	A09	Output Problem	11/05/2018	
									A0907	No Device Output	11/05/2018	
							F1905	Device Revision or Replacement	A09	Output Problem	11/05/2018	
									A0907	No Device Output	11/05/2018	
4879	2017	All other reportable incidents	35223	Endocardial/interventricular septal pacing lead	CT1684	Cardiac pacemakers and associated devices	E2401	Insufficient Information	A26	Insufficient Device Problem Information		
4880	2018	All other reportable incidents			CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A07	Electrical /Electronic Property Problem		
									A0708	Power Problem		
4881	2018	All other reportable incidents	47270	Cardiac resynchronization therapy implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A07	Electrical /Electronic Property Problem		
									A0709	Device Sensing Problem		
4882	2017	TBD	47270	Cardiac resynchronization therapy implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A07	Electrical /Electronic Property Problem		
									A0709	Device Sensing Problem		
4883	2018	All other reportable incidents	35223	Endocardial/interventricular septal pacing lead	CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A07	Electrical /Electronic Property Problem		
									A0709	Device Sensing Problem		
4884	2018	All other reportable incidents	47270	Cardiac resynchronization therapy implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A07	Electrical /Electronic Property Problem	17/08/2018	
									A0708	Power Problem	17/08/2018	
4885	2017	TBD			CT257	Implantable cardiac pacemakers	F05	Delay to Treatment/ Therapy	A12	Connection Problem		
									A1204	Failure to Disconnect		
4886	2017	All other reportable incidents			CT257	Implantable cardiac pacemakers	F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A07	Electrical /Electronic Property Problem		
									A0708	Power Problem		
									A13	Communication or Transmission Problem		
									A26	Insufficient Device Problem Information		
4887	2018	All other reportable incidents	35852	Single-chamber implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A07	Electrical /Electronic Property Problem		
									A0708	Power Problem		
							F1905	Device Revision or Replacement	A07	Electrical /Electronic Property Problem		
									A0708	Power Problem		
4888	2018	All other reportable incidents	35223	Endocardial/interventricular septal pacing lead	CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A26	Insufficient Device Problem Information		
4889	2018	All other reportable incidents	47263	Cardiac resynchronization therapy implantable pacemaker	CT1684	Cardiac pacemakers and associated devices	E2401	Insufficient Information	A11	Computer Software Problem		
									A1102	Application Program Problem		
					CT1685	Cardiac pacemakers	E2401	Insufficient Information	A11	Computer Software Problem		
									A1102	Application Program Problem		
					CT257	Implantable cardiac pacemakers	E2401	Insufficient Information	A11	Computer Software Problem		
									A1102	Application Program Problem		
					CT2866	Implantable cardiac electrical stimulators	E2401	Insufficient Information	A11	Computer Software Problem		
									A1102	Application Program Problem		
4890	2017	TBD	47270	Cardiac resynchronization therapy implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	F1001	Absence of Treatment	A07	Electrical /Electronic Property Problem		
									A0709	Device Sensing Problem		
4891	2017	All other reportable incidents	35852	Single-chamber implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A07	Electrical /Electronic Property Problem		
									A0709	Device Sensing Problem		
4892	2017	TBD	35852	Single-chamber implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A07	Electrical /Electronic Property Problem		
									A0709	Device Sensing Problem		
4893	2017	All other reportable incidents	47263	Cardiac resynchronization therapy implantable pacemaker	CT1684	Cardiac pacemakers and associated devices	F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A07	Electrical /Electronic Property Problem		
									A0709	Device Sensing Problem		
					CT1685	Cardiac pacemakers	F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A07	Electrical /Electronic Property Problem		
									A0709	Device Sensing Problem		
					CT257	Implantable cardiac pacemakers	F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A07	Electrical /Electronic Property Problem		
									A0709	Device Sensing Problem		
					CT2866	Implantable cardiac electrical stimulators	F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A07	Electrical /Electronic Property Problem		
									A0709	Device Sensing Problem		
4894	2018	All other reportable incidents	35223	Endocardial/interventricular septal pacing lead	CT1684	Cardiac pacemakers and associated devices	F1905	Device Revision or Replacement	A0104	Migration or Expulsion of Device		
4895	2018	All other reportable incidents	35223	Endocardial/interventricular septal pacing lead	CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A15	Activation, Positioning or Separation Problem		
									A150301	Separation Failure		
4896	2018	All other reportable incidents			CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A1104	Computer Operating System Problem		
4897	2017	TBD	47270	Cardiac resynchronization therapy implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A07	Electrical /Electronic Property Problem		
									A0708	Power Problem		
4898	2017	All other reportable incidents	47263	Cardiac resynchronization therapy implantable pacemaker	CT1684	Cardiac pacemakers and associated devices	F1905	Device Revision or Replacement	A07	Electrical /Electronic Property Problem		
									A0708	Power Problem		
					CT1685	Cardiac pacemakers	F1905	Device Revision or Replacement	A07	Electrical /Electronic Property Problem		

									A0708	Power Problem		
					CT257	Implantable cardiac pacemakers	F1905	Device Revision or Replacement	A07	Electrical /Electronic Property Problem		
									A0708	Power Problem		
					CT2866	Implantable cardiac electrical stimulators	F1905	Device Revision or Replacement	A07	Electrical /Electronic Property Problem		
									A0708	Power Problem		
4899	2017	All other reportable incidents	35223	Endocardial/interventricular septal pacing lead	CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A07	Electrical /Electronic Property Problem		
									A0709	Device Sensing Problem		
4900	2018	All other reportable incidents	61438	Cardiac pulse generator reader, home-use	CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A11	Computer Software Problem		
									A1102	Application Program Problem		
4901	2017	All other reportable incidents	35223	Endocardial/interventricular septal pacing lead	CT1684	Cardiac pacemakers and associated devices	F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A07	Electrical /Electronic Property Problem		
									A0709	Device Sensing Problem		
4902	2017	TBD			CT257	Implantable cardiac pacemakers			A26	Insufficient Device Problem Information		
4903	2017	All other reportable incidents	47270	Cardiac resynchronization therapy implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A07	Electrical /Electronic Property Problem		
									A0708	Power Problem		
4904	2018	All other reportable incidents			CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A09	Output Problem		
									A0907	No Device Output		
4905	2017	All other reportable incidents	35223	Endocardial/interventricular septal pacing lead	CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A04	Material Integrity Problem		
									A0405	Degraded		
4906	2017	TBD	47270	Cardiac resynchronization therapy implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A07	Electrical /Electronic Property Problem		
									A0708	Power Problem		
4907	2017	All other reportable incidents			CT257	Implantable cardiac pacemakers	F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A07	Electrical /Electronic Property Problem		
									A0709	Device Sensing Problem		
4908	2018	All other reportable incidents			CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A07	Electrical /Electronic Property Problem	05/09/2018	
									A0708	Power Problem	05/09/2018	
4909	2018	Death	60789	Intracardiac pacemaker	CT130	Catheters and associated devices	E2114	Perforation	A26	Insufficient Device Problem Information		
					CT145	Cardiovascular devices	E2114	Perforation	A26	Insufficient Device Problem Information		
					CT1684	Cardiac pacemakers and associated devices	E2114	Perforation	A26	Insufficient Device Problem Information		
					CT1685	Cardiac pacemakers	E2114	Perforation	A26	Insufficient Device Problem Information		
					CT233	Surgical	E2114	Perforation	A26	Insufficient Device Problem Information		
					CT2406	Implantable	E2114	Perforation	A26	Insufficient Device Problem Information		
					CT2407	Active implantable	E2114	Perforation	A26	Insufficient Device Problem Information		
					CT257	Implantable cardiac pacemakers	E2114	Perforation	A26	Insufficient Device Problem Information		
					CT2727	Active implant devices/ancillaries	E2114	Perforation	A26	Insufficient Device Problem Information		
					CT2750	Implantable cardiac electrical stimulators and associated devices	E2114	Perforation	A26	Insufficient Device Problem Information		
					CT2818	Catheterization/Cannulation/Fluid path devices	E2114	Perforation	A26	Insufficient Device Problem Information		
					CT2866	Implantable cardiac electrical stimulators	E2114	Perforation	A26	Insufficient Device Problem Information		
					CT2985	Active	E2114	Perforation	A26	Insufficient Device Problem Information		
					CT3001	Implantable metal	E2114	Perforation	A26	Insufficient Device Problem Information		
					CT3002	Implantable synthetic polymer	E2114	Perforation	A26	Insufficient Device Problem Information		
					CT3006	Implantable inorganic material	E2114	Perforation	A26	Insufficient Device Problem Information		
					CT314	Electrically powered	E2114	Perforation	A26	Insufficient Device Problem Information		
					CT321	Long-term surgical invasive	E2114	Perforation	A26	Insufficient Device Problem Information		
					CT981	Single use	E2114	Perforation	A26	Insufficient Device Problem Information		
					CT983	Surgical invasive	E2114	Perforation	A26	Insufficient Device Problem Information		
4910	2018	All other reportable incidents	35223	Endocardial/interventricular septal pacing lead	CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A05	Mechanical Problem	10/09/2018	
									A0510	Retraction Problem	10/09/2018	
									A15	Activation, Positioning or Separation Problem	10/09/2018	
									A150203	Difficult or Delayed Positioning	10/09/2018	
4911	2018	All other reportable incidents			CT257	Implantable cardiac pacemakers	F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A07	Electrical /Electronic Property Problem	10/05/2018	
									A0708	Power Problem	10/05/2018	
4912	2017	Unanticipated serious deterioration in state of health	47270	Cardiac resynchronization therapy implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	F1905	Device Revision or Replacement	A07	Electrical /Electronic Property Problem	23/02/2017	
									A0708	Power Problem	23/02/2017	
4913	2018	All other reportable incidents			CT257	Implantable cardiac pacemakers	E0402	Hypersensitivity/Allergic Reaction	A11	Computer Software Problem		
									A1102	Application Program Problem		
4914	2018	All other reportable incidents			CT257	Implantable cardiac pacemakers	F1905	Device Revision or Replacement	A07	Electrical /Electronic Property Problem	04/07/2018	
									A0708	Power Problem	04/07/2018	
4915	2017	TBD	47270	Cardiac resynchronization therapy implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	F1905	Device Revision or Replacement	A07	Electrical /Electronic Property Problem		
									A0708	Power Problem		
4916	2018	All other reportable incidents			CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A09	Output Problem		
									A0907	No Device Output		
4917	2018	All other reportable incidents	35223	Endocardial/interventricular septal pacing lead	CT1684	Cardiac pacemakers and associated devices	F1905	Device Revision or Replacement	A0104	Migration or Expulsion of Device	09/05/2018	
									A07	Electrical /Electronic Property Problem	09/05/2018	
									A0709	Device Sensing Problem	09/05/2018	
4918	2017	All other reportable incidents			CT257	Implantable cardiac pacemakers	F1905	Device Revision or Replacement	A09	Output Problem	04/07/2017	
									A0907	No Device Output	04/07/2017	
4919	2017	All other reportable incidents			CT1684	Cardiac pacemakers and associated devices	F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A09	Output Problem		
									A0908	Incorrect, Inadequate or Imprecise Result or Readings		
4920	2017	TBD	35852	Single-chamber implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A09	Output Problem		
									A090404	Output Above Specifications		
4921	2017	TBD	47270	Cardiac resynchronization therapy implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	F14	Prolonged Episode of Care	A26	Insufficient Device Problem Information		
4922	2017	TBD	35852	Single-chamber implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A1304	Telemetry Discrepancy		
									A26	Insufficient Device Problem Information		
4923	2017	TBD	35852	Single-chamber implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A07	Electrical /Electronic Property Problem		
									A0709	Device Sensing Problem		
4924	2017	TBD	47270	Cardiac resynchronization therapy implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A16	Protective Measures Problem		
									A1601	Device Alarm System		
4925	2017	All other reportable incidents			CT257	Implantable cardiac pacemakers	F1905	Device Revision or Replacement	A07	Electrical /Electronic Property Problem		
									A0708	Power Problem		
4926	2018	All other reportable incidents			CT257	Implantable cardiac pacemakers	F1905	Device Revision or Replacement	A07	Electrical /Electronic Property Problem	02/10/2018	
									A0708	Power Problem	02/10/2018	
4927	2018	All other reportable incidents	35223	Endocardial/interventricular septal pacing lead	CT1684	Cardiac pacemakers and associated devices	F1905	Device Revision or Replacement	A04	Material Integrity Problem	21/09/2018	
									A0405	Degraded	21/09/2018	
4928	2017	TBD	35852	Single-chamber implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A07	Electrical /Electronic Property Problem		
									A0709	Device Sensing Problem		
4929	2018	All other reportable incidents	35852	Single-chamber implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A07	Electrical /Electronic Property Problem		

									A0709	Device Sensing Problem		
4930	2018	All other reportable incidents			CT257	Implantable cardiac pacemakers	F1905	Device Revision or Replacement	A07	Electrical /Electronic Property Problem		
									A0709	Device Sensing Problem		
4931	2017	All other reportable incidents			CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A07	Electrical /Electronic Property Problem		
									A0708	Power Problem		
4932	2017	All other reportable incidents	47270	Cardiac resynchronization therapy implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	F1905	Device Revision or Replacement	A07	Electrical /Electronic Property Problem	08/12/2017	
									A0708	Power Problem	08/12/2017	
4933	2017	TBD	35852	Single-chamber implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A07	Electrical /Electronic Property Problem		
									A0709	Device Sensing Problem		
4934	2018	All other reportable incidents			CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A07	Electrical /Electronic Property Problem		
									A0708	Power Problem		
4935	2018	All other reportable incidents	47270	Cardiac resynchronization therapy implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A07	Electrical /Electronic Property Problem		
									A0709	Device Sensing Problem		
4936	2018	All other reportable incidents	35852	Single-chamber implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A07	Electrical /Electronic Property Problem		
									A0709	Device Sensing Problem		
4937	2018	All other reportable incidents	35852	Single-chamber implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A07	Electrical /Electronic Property Problem		
									A0709	Device Sensing Problem		
4938	2017	All other reportable incidents	47270	Cardiac resynchronization therapy implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A07	Electrical /Electronic Property Problem		
									A0709	Device Sensing Problem		
4939	2017	TBD	35852	Single-chamber implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A07	Electrical /Electronic Property Problem		
									A0709	Device Sensing Problem		
4940	2017	All other reportable incidents			CT257	Implantable cardiac pacemakers	F1905	Device Revision or Replacement	A07	Electrical /Electronic Property Problem		
									A0709	Device Sensing Problem		
4941	2017	All other reportable incidents	47270	Cardiac resynchronization therapy implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	F21	Unexpected Deterioration	A26	Insufficient Device Problem Information		
4942	2018	All other reportable incidents			CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A07	Electrical /Electronic Property Problem		
									A0709	Device Sensing Problem		
4943	2017	TBD	35852	Single-chamber implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A07	Electrical /Electronic Property Problem		
									A0709	Device Sensing Problem		
4944	2018	All other reportable incidents	35223	Endocardial/interventricular septal pacing lead	CT1684	Cardiac pacemakers and associated devices	F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A07	Electrical /Electronic Property Problem		
									A0709	Device Sensing Problem		
4945	2017	TBD	35223	Endocardial/interventricular septal pacing lead	CT1684	Cardiac pacemakers and associated devices	F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A26	Insufficient Device Problem Information		
4946	2018	All other reportable incidents			CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A07	Electrical /Electronic Property Problem		
									A0708	Power Problem		
4947	2017	All other reportable incidents			CT1684	Cardiac pacemakers and associated devices	E0402	Hypersensitivity/Allergic Reaction	A1104	Computer Operating System Problem		
4948	2017	TBD	35223	Endocardial/interventricular septal pacing lead	CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A26	Insufficient Device Problem Information		
4949	2017	TBD	47270	Cardiac resynchronization therapy implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A07	Electrical /Electronic Property Problem		
									A0708	Power Problem		
4950	2017	TBD	47270	Cardiac resynchronization therapy implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A07	Electrical /Electronic Property Problem		
									A0709	Device Sensing Problem		
4951	2018	All other reportable incidents	61438	Cardiac pulse generator reader, home-use	CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A11	Computer Software Problem		
									A1102	Application Program Problem		
4952	2018	All other reportable incidents	35223	Endocardial/interventricular septal pacing lead	CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A12	Connection Problem		
4953	2017	TBD	47270	Cardiac resynchronization therapy implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A07	Electrical /Electronic Property Problem		
									A0708	Power Problem		
4954	2017	All other reportable incidents	35223	Endocardial/interventricular septal pacing lead	CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A15	Activation, Positioning or Separation Problem		
									A150203	Difficult or Delayed Positioning		
4955	2017	All other reportable incidents	35223	Endocardial/interventricular septal pacing lead	CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A15	Activation, Positioning or Separation Problem	21/11/2017	
									A150203	Difficult or Delayed Positioning	21/11/2017	
4956	2018	All other reportable incidents	47270	Cardiac resynchronization therapy implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	F1905	Device Revision or Replacement	A07	Electrical /Electronic Property Problem		
									A0708	Power Problem		
4957	2017	All other reportable incidents			CT257	Implantable cardiac pacemakers	F1905	Device Revision or Replacement	A07	Electrical /Electronic Property Problem	26/01/2017	
									A0708	Power Problem	26/01/2017	
4958	2017	TBD			CT257	Implantable cardiac pacemakers	F14	Prolonged Episode of Care	A05	Mechanical Problem		
									A0501	Detachment of Device or Device Component		
							F1905	Device Revision or Replacement	A05	Mechanical Problem		
									A0501	Detachment of Device or Device Component		
4959	2017	All other reportable incidents			CT257	Implantable cardiac pacemakers	F1905	Device Revision or Replacement	A07	Electrical /Electronic Property Problem	24/11/2017	
									A0708	Power Problem	24/11/2017	
4960	2018	All other reportable incidents	47270	Cardiac resynchronization therapy implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A07	Electrical /Electronic Property Problem		
									A0709	Device Sensing Problem		
4961	2018	Unanticipated serious deterioration in state of health	35223	Endocardial/interventricular septal pacing lead	CT1684	Cardiac pacemakers and associated devices	F1905	Device Revision or Replacement	A07	Electrical /Electronic Property Problem		
									A0709	Device Sensing Problem		
4962	2018	Unanticipated serious deterioration in state of health	35223	Endocardial/interventricular septal pacing lead	CT1684	Cardiac pacemakers and associated devices	E1603	Bone Fracture(s)	A07	Electrical /Electronic Property Problem		
									A0709	Device Sensing Problem		
							E2007	Fall	A07	Electrical /Electronic Property Problem		
									A0709	Device Sensing Problem		
4963	2018	Unanticipated serious deterioration in state of health			CT257	Implantable cardiac pacemakers	E1603	Bone Fracture(s)	A07	Electrical /Electronic Property Problem	27/09/2018	
									A0709	Device Sensing Problem	27/09/2018	
							E2007	Fall	A07	Electrical /Electronic Property Problem	27/09/2018	
									A0709	Device Sensing Problem	27/09/2018	
							F14	Prolonged Episode of Care	A07	Electrical /Electronic Property Problem	27/09/2018	
									A0709	Device Sensing Problem	27/09/2018	
							F1905	Device Revision or Replacement	A07	Electrical /Electronic Property Problem	27/09/2018	
									A0709	Device Sensing Problem	27/09/2018	
4964	2018	All other reportable incidents	47270	Cardiac resynchronization therapy implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A07	Electrical /Electronic Property Problem		
									A0709	Device Sensing Problem		
4965	2018	All other reportable incidents	35223	Endocardial/interventricular septal pacing lead	CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A0203	Defective Device		
									A26	Insufficient Device Problem Information		
4966	2017	All other reportable incidents			CT257	Implantable cardiac pacemakers	F1905	Device Revision or Replacement	A1304	Telemetry Discrepancy	12/06/2017	
									A26	Insufficient Device Problem Information	12/06/2017	
4967	2017	All other reportable incidents	47270	Cardiac resynchronization therapy implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	F1905	Device Revision or Replacement	A07	Electrical /Electronic Property Problem	10/07/2017	
									A0708	Power Problem	10/07/2017	
4968	2017	All other reportable incidents	47270	Cardiac resynchronization therapy implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	F1905	Device Revision or Replacement	A07	Electrical /Electronic Property Problem	04/07/2017	
									A0708	Power Problem	04/07/2017	
4969	2017	All other reportable incidents			CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A09	Output Problem		

									A0907	No Device Output		
4970	2017	TBD	35223	Endocardial/interventricular septal pacing lead	CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A07	Electrical /Electronic Property Problem		
									A0709	Device Sensing Problem		
4971	2017	All other reportable incidents	47270	Cardiac resynchronization therapy implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A07	Electrical /Electronic Property Problem		
									A0708	Power Problem		
									A0709	Device Sensing Problem		
4972	2017	All other reportable incidents	35223	Endocardial/interventricular septal pacing lead	CT1684	Cardiac pacemakers and associated devices	F1905	Device Revision or Replacement	A26	Insufficient Device Problem Information		
4973	2017	TBD			CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A07	Electrical /Electronic Property Problem		
									A0709	Device Sensing Problem		
4974	2017	TBD			CT257	Implantable cardiac pacemakers	F1905	Device Revision or Replacement	A13	Communication or Transmission Problem		
									A26	Insufficient Device Problem Information		
4975	2017	All other reportable incidents	35223	Endocardial/interventricular septal pacing lead	CT1684	Cardiac pacemakers and associated devices	F14	Prolonged Episode of Care	A07	Electrical /Electronic Property Problem	07/12/2017	
									A0709	Device Sensing Problem	07/12/2017	
4976	2018	All other reportable incidents			CT257	Implantable cardiac pacemakers	F1905	Device Revision or Replacement	A07	Electrical /Electronic Property Problem	03/09/2018	
									A0709	Device Sensing Problem	03/09/2018	
4977	2017	All other reportable incidents	35852	Single-chamber implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E2104	Electric Shock	A07	Electrical /Electronic Property Problem		
									A0709	Device Sensing Problem		
4978	2018	All other reportable incidents	35223	Endocardial/interventricular septal pacing lead	CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A05	Mechanical Problem		
									A0510	Retraction Problem		
4979	2018	All other reportable incidents	47270	Cardiac resynchronization therapy implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A11	Computer Software Problem		
									A1102	Application Program Problem		
4980	2017	All other reportable incidents			CT257	Implantable cardiac pacemakers	F1905	Device Revision or Replacement	A01	Patient Device Interaction Problem		
									A17	Compatibility Problem		
4981	2018	All other reportable incidents	47270	Cardiac resynchronization therapy implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A11	Computer Software Problem		
									A1102	Application Program Problem		
4982	2018	All other reportable incidents	35852	Single-chamber implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A07	Electrical /Electronic Property Problem		
									A0709	Device Sensing Problem		
4983	2018	All other reportable incidents	35852	Single-chamber implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A07	Electrical /Electronic Property Problem		
									A0709	Device Sensing Problem		
4984	2018	All other reportable incidents	35223	Endocardial/interventricular septal pacing lead	CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A04	Material Integrity Problem		
									A0413	Material Separation		
4985	2018	All other reportable incidents	35852	Single-chamber implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A07	Electrical /Electronic Property Problem		
									A0709	Device Sensing Problem		
4986	2018	All other reportable incidents			CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A09	Output Problem		
									A0907	No Device Output		
4987	2017	TBD	47270	Cardiac resynchronization therapy implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A07	Electrical /Electronic Property Problem		
									A0708	Power Problem		
4988	2018	All other reportable incidents			CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A1104	Computer Operating System Problem		
4989	2018	All other reportable incidents			CT1684	Cardiac pacemakers and associated devices	F05	Delay to Treatment/ Therapy	A07	Electrical /Electronic Property Problem		
									A0708	Power Problem		
4990	2017	All other reportable incidents	47270	Cardiac resynchronization therapy implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	F1905	Device Revision or Replacement	A07	Electrical /Electronic Property Problem	03/02/2017	
									A0708	Power Problem	03/02/2017	
4991	2017	All other reportable incidents	47263	Cardiac resynchronization therapy implantable pacemaker	CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A26	Insufficient Device Problem Information	31/01/2016	
					CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A26	Insufficient Device Problem Information	31/01/2016	
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A26	Insufficient Device Problem Information	31/01/2016	
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A26	Insufficient Device Problem Information	31/01/2016	
4992	2017	TBD			CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A1104	Computer Operating System Problem		
4993	2017	All other reportable incidents	47270	Cardiac resynchronization therapy implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	F1905	Device Revision or Replacement	A07	Electrical /Electronic Property Problem		
									A0708	Power Problem		
4994	2017	All other reportable incidents	35223	Endocardial/interventricular septal pacing lead	CT1684	Cardiac pacemakers and associated devices	F1905	Device Revision or Replacement	A07	Electrical /Electronic Property Problem		
									A0709	Device Sensing Problem		
4995	2017	TBD			CT257	Implantable cardiac pacemakers	F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A13	Communication or Transmission Problem		
									A26	Insufficient Device Problem Information		
4996	2017	TBD			CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A07	Electrical /Electronic Property Problem		
									A0708	Power Problem		
4997	2017	TBD			CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A07	Electrical /Electronic Property Problem		
									A0708	Power Problem		
4998	2018	All other reportable incidents	35223	Endocardial/interventricular septal pacing lead	CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A05	Mechanical Problem		
									A0510	Retraction Problem		
4999	2018	All other reportable incidents	35223	Endocardial/interventricular septal pacing lead	CT1684	Cardiac pacemakers and associated devices	F1905	Device Revision or Replacement	A05	Mechanical Problem		
									A0510	Retraction Problem		
									A07	Electrical /Electronic Property Problem		
									A0709	Device Sensing Problem		
5000	2017	All other reportable incidents	35852	Single-chamber implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A07	Electrical /Electronic Property Problem		
									A0709	Device Sensing Problem		
5001	2017	TBD	35852	Single-chamber implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A07	Electrical /Electronic Property Problem		
									A0709	Device Sensing Problem		
5002	2017	All other reportable incidents	47270	Cardiac resynchronization therapy implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A07	Electrical /Electronic Property Problem		
									A0709	Device Sensing Problem		
5003	2017	All other reportable incidents	35852	Single-chamber implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A07	Electrical /Electronic Property Problem		
									A0709	Device Sensing Problem		
5004	2017	All other reportable incidents	35852	Single-chamber implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A07	Electrical /Electronic Property Problem		
									A0709	Device Sensing Problem		
5005	2017	TBD	35852	Single-chamber implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A09	Output Problem		
									A090404	Output Above Specifications		
5006	2017	TBD	35852	Single-chamber implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A07	Electrical /Electronic Property Problem		
									A0709	Device Sensing Problem		
5007	2017	TBD	35852	Single-chamber implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A07	Electrical /Electronic Property Problem		
									A0709	Device Sensing Problem		
5008	2017	All other reportable incidents	35852	Single-chamber implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A07	Electrical /Electronic Property Problem		
									A0709	Device Sensing Problem		
									A09	Output Problem		
									A090404	Output Above Specifications		
5009	2017	TBD	35852	Single-chamber implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A07	Electrical /Electronic Property Problem		
									A0709	Device Sensing Problem		
5010	2017	TBD	35223	Endocardial/interventricular septal pacing lead	CT1684	Cardiac pacemakers and associated devices	E2401	Insufficient Information	A26	Insufficient Device Problem Information		

5011	2018	All other reportable incidents			CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A07	Electrical /Electronic Property Problem	28/11/2017	
									A0708	Power Problem	28/11/2017	
5012	2018	All other reportable incidents	47270	Cardiac resynchronization therapy implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	F1905	Device Revision or Replacement	A07	Electrical /Electronic Property Problem	25/04/2018	
									A0708	Power Problem	25/04/2018	
5013	2017	TBD	35223	Endocardial/interventricular septal pacing lead	CT1684	Cardiac pacemakers and associated devices	F21	Unexpected Deterioration	A26	Insufficient Device Problem Information		
5014	2018	All other reportable incidents			CT1684	Cardiac pacemakers and associated devices	F05	Delay to Treatment/ Therapy	A09	Output Problem		
									A0907	No Device Output		
									A26	Insufficient Device Problem Information		
5015	2017	All other reportable incidents	60789	Intracardiac pacemaker	CT130	Catheters and associated devices	E1906	Unspecified Infection	A26	Insufficient Device Problem Information		
							E2114	Perforation	A26	Insufficient Device Problem Information		
					CT145	Cardiovascular devices	E1906	Unspecified Infection	A26	Insufficient Device Problem Information		
							E2114	Perforation	A26	Insufficient Device Problem Information		
					CT1684	Cardiac pacemakers and associated devices	E1906	Unspecified Infection	A26	Insufficient Device Problem Information		
							E2114	Perforation	A26	Insufficient Device Problem Information		
					CT1685	Cardiac pacemakers	E1906	Unspecified Infection	A26	Insufficient Device Problem Information		
							E2114	Perforation	A26	Insufficient Device Problem Information		
					CT233	Surgical	E1906	Unspecified Infection	A26	Insufficient Device Problem Information		
							E2114	Perforation	A26	Insufficient Device Problem Information		
					CT2406	Implantable	E1906	Unspecified Infection	A26	Insufficient Device Problem Information		
							E2114	Perforation	A26	Insufficient Device Problem Information		
					CT2407	Active implantable	E1906	Unspecified Infection	A26	Insufficient Device Problem Information		
							E2114	Perforation	A26	Insufficient Device Problem Information		
					CT257	Implantable cardiac pacemakers	E1906	Unspecified Infection	A26	Insufficient Device Problem Information		
							E2114	Perforation	A26	Insufficient Device Problem Information		
					CT2727	Active implant devices/ancillaries	E1906	Unspecified Infection	A26	Insufficient Device Problem Information		
							E2114	Perforation	A26	Insufficient Device Problem Information		
					CT2750	Implantable cardiac electrical stimulators and associated devices	E1906	Unspecified Infection	A26	Insufficient Device Problem Information		
							E2114	Perforation	A26	Insufficient Device Problem Information		
					CT2818	Catheterization/Cannulation/Fluid path devices	E1906	Unspecified Infection	A26	Insufficient Device Problem Information		
							E2114	Perforation	A26	Insufficient Device Problem Information		
					CT2866	Implantable cardiac electrical stimulators	E1906	Unspecified Infection	A26	Insufficient Device Problem Information		
							E2114	Perforation	A26	Insufficient Device Problem Information		
					CT2985	Active	E1906	Unspecified Infection	A26	Insufficient Device Problem Information		
							E2114	Perforation	A26	Insufficient Device Problem Information		
					CT3001	Implantable metal	E1906	Unspecified Infection	A26	Insufficient Device Problem Information		
							E2114	Perforation	A26	Insufficient Device Problem Information		
					CT3002	Implantable synthetic polymer	E1906	Unspecified Infection	A26	Insufficient Device Problem Information		
							E2114	Perforation	A26	Insufficient Device Problem Information		
					CT3006	Implantable inorganic material	E1906	Unspecified Infection	A26	Insufficient Device Problem Information		
							E2114	Perforation	A26	Insufficient Device Problem Information		
					CT314	Electrically powered	E1906	Unspecified Infection	A26	Insufficient Device Problem Information		
							E2114	Perforation	A26	Insufficient Device Problem Information		
					CT321	Long-term surgical invasive	E1906	Unspecified Infection	A26	Insufficient Device Problem Information		
							E2114	Perforation	A26	Insufficient Device Problem Information		
					CT981	Single use	E1906	Unspecified Infection	A26	Insufficient Device Problem Information		
							E2114	Perforation	A26	Insufficient Device Problem Information		
					CT983	Surgical invasive	E1906	Unspecified Infection	A26	Insufficient Device Problem Information		
							E2114	Perforation	A26	Insufficient Device Problem Information		
5016	2017	All other reportable incidents	47270	Cardiac resynchronization therapy implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	F1905	Device Revision or Replacement	A07	Electrical /Electronic Property Problem	07/12/2017	
									A0708	Power Problem	07/12/2017	
5017	2018	All other reportable incidents	35852	Single-chamber implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E2104	Electric Shock	A07	Electrical /Electronic Property Problem		
									A0709	Device Sensing Problem		
							F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A07	Electrical /Electronic Property Problem		
									A0709	Device Sensing Problem		
5018	2018	All other reportable incidents	35223	Endocardial/interventricular septal pacing lead	CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A05	Mechanical Problem		
									A0510	Retraction Problem		
5019	2017	Death	47270	Cardiac resynchronization therapy implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A09	Output Problem		
									A0907	No Device Output		
5020	2018	All other reportable incidents			CT257	Implantable cardiac pacemakers	E2007	Fall	A07	Electrical /Electronic Property Problem	24/04/2018	
									A0709	Device Sensing Problem	24/04/2018	
									A13	Communication or Transmission Problem	24/04/2018	
									A26	Insufficient Device Problem Information	24/04/2018	
							F14	Prolonged Episode of Care	A07	Electrical /Electronic Property Problem	24/04/2018	
									A0709	Device Sensing Problem	24/04/2018	
									A13	Communication or Transmission Problem	24/04/2018	
									A26	Insufficient Device Problem Information	24/04/2018	
							F1905	Device Revision or Replacement	A07	Electrical /Electronic Property Problem	24/04/2018	
									A0709	Device Sensing Problem	24/04/2018	
									A13	Communication or Transmission Problem	24/04/2018	
									A26	Insufficient Device Problem Information	24/04/2018	
5021	2017	All other reportable incidents			CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A07	Electrical /Electronic Property Problem		
									A0709	Device Sensing Problem		
5022	2018	All other reportable incidents	35852	Single-chamber implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	F1905	Device Revision or Replacement	A07	Electrical /Electronic Property Problem	15/10/2018	
									A0708	Power Problem	15/10/2018	
5023	2017	TBD	47263	Cardiac resynchronization therapy implantable pacemaker	CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A07	Electrical /Electronic Property Problem		
									A0709	Device Sensing Problem		
									A09	Output Problem		
									A0908	Incorrect, Inadequate or Imprecise Result or Readings		
					CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A07	Electrical /Electronic Property Problem		
									A0709	Device Sensing Problem		
									A09	Output Problem		
									A0908	Incorrect, Inadequate or Imprecise Result or Readings		
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A07	Electrical /Electronic Property Problem		
									A0709	Device Sensing Problem		
									A09	Output Problem		

									A0908	Incorrect, Inadequate or Imprecise Result or Readings		
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A07	Electrical /Electronic Property Problem		
									A0709	Device Sensing Problem		
									A09	Output Problem		
									A0908	Incorrect, Inadequate or Imprecise Result or Readings		
5024	2017	TBD	47270	Cardiac resynchronization therapy implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A07	Electrical /Electronic Property Problem		
									A0708	Power Problem		
5025	2018	All other reportable incidents	47270	Cardiac resynchronization therapy implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	F1905	Device Revision or Replacement	A07	Electrical /Electronic Property Problem	26/02/2018	
									A0708	Power Problem	26/02/2018	
5026	2017	All other reportable incidents	47270	Cardiac resynchronization therapy implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	F1905	Device Revision or Replacement	A07	Electrical /Electronic Property Problem	21/12/2017	
									A0708	Power Problem	21/12/2017	
5027	2018	All other reportable incidents	35223	Endocardial/interventricular septal pacing lead	CT1684	Cardiac pacemakers and associated devices	F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A07	Electrical /Electronic Property Problem		
									A0709	Device Sensing Problem		
							F1905	Device Revision or Replacement	A07	Electrical /Electronic Property Problem		
									A0709	Device Sensing Problem		
5028	2017	All other reportable incidents	47270	Cardiac resynchronization therapy implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A07	Electrical /Electronic Property Problem	06/12/2017	
									A0708	Power Problem	06/12/2017	
							F1905	Device Revision or Replacement	A07	Electrical /Electronic Property Problem	06/12/2017	
									A0708	Power Problem	06/12/2017	
5029	2017	TBD			CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A07	Electrical /Electronic Property Problem		
									A0708	Power Problem		
5030	2018	All other reportable incidents	47270	Cardiac resynchronization therapy implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A07	Electrical /Electronic Property Problem		
									A0709	Device Sensing Problem		
									A070909	Oversensing		
									A0722	Impedance Problem		
									A072201	High Impedance		
5031	2018	All other reportable incidents	35223	Endocardial/interventricular septal pacing lead	CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A07	Electrical /Electronic Property Problem		
									A0709	Device Sensing Problem		
5032	2018	All other reportable incidents	35223	Endocardial/interventricular septal pacing lead	CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A07	Electrical /Electronic Property Problem		
									A0709	Device Sensing Problem		
5033	2017	All other reportable incidents	47270	Cardiac resynchronization therapy implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	F1905	Device Revision or Replacement	A07	Electrical /Electronic Property Problem	07/12/2017	
									A0708	Power Problem	07/12/2017	
5034	2017	All other reportable incidents	47270	Cardiac resynchronization therapy implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	F1905	Device Revision or Replacement	A07	Electrical /Electronic Property Problem	07/08/2017	
									A0708	Power Problem	07/08/2017	
5035	2017	All other reportable incidents			CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A11	Computer Software Problem		
									A1102	Application Program Problem		
5036	2018	Unanticipated serious deterioration in state of health			CT257	Implantable cardiac pacemakers	E2401	Insufficient Information	A26	Insufficient Device Problem Information		
5037	2018	All other reportable incidents	47270	Cardiac resynchronization therapy implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A07	Electrical /Electronic Property Problem		
									A0709	Device Sensing Problem		
5038	2018	All other reportable incidents			CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A07	Electrical /Electronic Property Problem	18/04/2018	
									A0708	Power Problem	18/04/2018	
5039	2017	All other reportable incidents	35223	Endocardial/interventricular septal pacing lead	CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A15	Activation, Positioning or Separation Problem		
									A150203	Difficult or Delayed Positioning		
5040	2017	TBD	35223	Endocardial/interventricular septal pacing lead	CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A15	Activation, Positioning or Separation Problem		
									A150203	Difficult or Delayed Positioning		
5041	2017	TBD	35223	Endocardial/interventricular septal pacing lead	CT1684	Cardiac pacemakers and associated devices	F05	Delay to Treatment/ Therapy	A07	Electrical /Electronic Property Problem		
									A0709	Device Sensing Problem		
5042	2017	TBD	35223	Endocardial/interventricular septal pacing lead	CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A07	Electrical /Electronic Property Problem		
									A0709	Device Sensing Problem		
5043	2017	TBD	35223	Endocardial/interventricular septal pacing lead	CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A15	Activation, Positioning or Separation Problem		
									A150203	Difficult or Delayed Positioning		
									A150301	Separation Failure		
5044	2017	TBD	35223	Endocardial/interventricular septal pacing lead	CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A05	Mechanical Problem		
									A0510	Retraction Problem		
5045	2017	TBD			CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A07	Electrical /Electronic Property Problem		
									A0709	Device Sensing Problem		
5046	2017	TBD	35852	Single-chamber implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A07	Electrical /Electronic Property Problem		
									A0709	Device Sensing Problem		
5047	2017	TBD	35223	Endocardial/interventricular septal pacing lead	CT1684	Cardiac pacemakers and associated devices	F1905	Device Revision or Replacement	A26	Insufficient Device Problem Information		
5048	2018	All other reportable incidents	35852	Single-chamber implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A07	Electrical /Electronic Property Problem		
									A0709	Device Sensing Problem		
5049	2017	All other reportable incidents	35223	Endocardial/interventricular septal pacing lead	CT1684	Cardiac pacemakers and associated devices	E2401	Insufficient Information	A26	Insufficient Device Problem Information		
5050	2018	All other reportable incidents	35852	Single-chamber implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E2104	Electric Shock	A07	Electrical /Electronic Property Problem		
									A0709	Device Sensing Problem		
5051	2018	All other reportable incidents	35223	Endocardial/interventricular septal pacing lead	CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A07	Electrical /Electronic Property Problem		
									A0709	Device Sensing Problem		
5052	2017	All other reportable incidents	47270	Cardiac resynchronization therapy implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A07	Electrical /Electronic Property Problem		
									A0708	Power Problem		
5053	2017	All other reportable incidents	47270	Cardiac resynchronization therapy implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A07	Electrical /Electronic Property Problem		
									A0708	Power Problem		
5054	2017	TBD	47270	Cardiac resynchronization therapy implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A07	Electrical /Electronic Property Problem		
									A0708	Power Problem		
5055	2017	All other reportable incidents	35223	Endocardial/interventricular septal pacing lead	CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A15	Activation, Positioning or Separation Problem		
									A150203	Difficult or Delayed Positioning		
5056	2017	All other reportable incidents	47270	Cardiac resynchronization therapy implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A07	Electrical /Electronic Property Problem		
									A0709	Device Sensing Problem		
5057	2017	All other reportable incidents			CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A11	Computer Software Problem		
									A1102	Application Program Problem		
5058	2017	TBD	35852	Single-chamber implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A07	Electrical /Electronic Property Problem		
									A0709	Device Sensing Problem		
5059	2017	All other reportable incidents	47270	Cardiac resynchronization therapy implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	F1905	Device Revision or Replacement	A13	Communication or Transmission Problem	10/02/2017	
									A26	Insufficient Device Problem Information	10/02/2017	
5060	2017	All other reportable incidents	47270	Cardiac resynchronization therapy implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	F1905	Device Revision or Replacement	A07	Electrical /Electronic Property Problem	06/12/2017	
									A0708	Power Problem	06/12/2017	

5061	2017	TBD	47270	Cardiac resynchronization therapy implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A07	Electrical /Electronic Property Problem		
									A0708	Power Problem		
5062	2017	All other reportable incidents	35223	Endocardial/interventricular septal pacing lead	CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A15	Activation, Positioning or Separation Problem		
									A150203	Difficult or Delayed Positioning		
5063	2017	All other reportable incidents			CT257	Implantable cardiac pacemakers	F1905	Device Revision or Replacement	A07	Electrical /Electronic Property Problem	19/07/2017	
									A0708	Power Problem	19/07/2017	
5064	2018	All other reportable incidents			CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A09	Output Problem		
									A0907	No Device Output		
5065	2017	All other reportable incidents			CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A09	Output Problem		
									A0907	No Device Output		
5066	2017	All other reportable incidents	47270	Cardiac resynchronization therapy implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	F1905	Device Revision or Replacement	A07	Electrical /Electronic Property Problem	06/12/2017	
									A0708	Power Problem	06/12/2017	
5067	2018	All other reportable incidents	47270	Cardiac resynchronization therapy implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A07	Electrical /Electronic Property Problem		
									A0708	Power Problem		
5068	2017	All other reportable incidents	47270	Cardiac resynchronization therapy implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A1104	Computer Operating System Problem		
5069	2017	All other reportable incidents	47270	Cardiac resynchronization therapy implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	F1905	Device Revision or Replacement	A07	Electrical /Electronic Property Problem	08/08/2017	
									A0708	Power Problem	08/08/2017	
5070	2018	All other reportable incidents	35852	Single-chamber implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E1906	Unspecified Infection	A01	Patient Device Interaction Problem		
									A17	Compatibility Problem		
5071	2017	All other reportable incidents	47270	Cardiac resynchronization therapy implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	F1905	Device Revision or Replacement	A07	Electrical /Electronic Property Problem	12/10/2017	
									A0708	Power Problem	12/10/2017	
5072	2017	All other reportable incidents			CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A07	Electrical /Electronic Property Problem		
									A0709	Device Sensing Problem		
5073	2017	TBD	47270	Cardiac resynchronization therapy implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A09	Output Problem		
									A0907	No Device Output		
5074	2018	All other reportable incidents	47270	Cardiac resynchronization therapy implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A07	Electrical /Electronic Property Problem		
									A0709	Device Sensing Problem		
5075	2017	Death			CT257	Implantable cardiac pacemakers	F1905	Device Revision or Replacement	A26	Insufficient Device Problem Information	18/12/2017	
5076	2018	All other reportable incidents	47270	Cardiac resynchronization therapy implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	F1905	Device Revision or Replacement	A07	Electrical /Electronic Property Problem		
									A0708	Power Problem		
5077	2017	Death			CT1684	Cardiac pacemakers and associated devices	F21	Unexpected Deterioration	A09	Output Problem		
									A0907	No Device Output		
5078	2018	All other reportable incidents	35223	Endocardial/interventricular septal pacing lead	CT1684	Cardiac pacemakers and associated devices	F1905	Device Revision or Replacement	A07	Electrical /Electronic Property Problem		
									A0709	Device Sensing Problem		
5079	2018	All other reportable incidents	35223	Endocardial/interventricular septal pacing lead	CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A07	Electrical /Electronic Property Problem		
									A0709	Device Sensing Problem		
5080	2018	All other reportable incidents	47270	Cardiac resynchronization therapy implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A07	Electrical /Electronic Property Problem		
									A0709	Device Sensing Problem		
5081	2018	All other reportable incidents			CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A07	Electrical /Electronic Property Problem		
									A0708	Power Problem		
5082	2017	TBD	35223	Endocardial/interventricular septal pacing lead	CT1684	Cardiac pacemakers and associated devices	E0402	Hypersensitivity/Allergic Reaction	A26	Insufficient Device Problem Information		
5083	2018	All other reportable incidents			CT1684	Cardiac pacemakers and associated devices	F05	Delay to Treatment/ Therapy	A09	Output Problem		
									A0907	No Device Output		
5084	2017	TBD	47270	Cardiac resynchronization therapy implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A07	Electrical /Electronic Property Problem		
									A0708	Power Problem		
5085	2018	All other reportable incidents			CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A07	Electrical /Electronic Property Problem		
									A0709	Device Sensing Problem		
5086	2018	All other reportable incidents			CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A07	Electrical /Electronic Property Problem		
									A0709	Device Sensing Problem		
5087	2018	All other reportable incidents	47270	Cardiac resynchronization therapy implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A07	Electrical /Electronic Property Problem	23/05/2018	
									A0708	Power Problem	23/05/2018	
5088	2017	All other reportable incidents			CT257	Implantable cardiac pacemakers	F1905	Device Revision or Replacement	A07	Electrical /Electronic Property Problem		
									A0708	Power Problem		
5089	2017	TBD	35223	Endocardial/interventricular septal pacing lead	CT1684	Cardiac pacemakers and associated devices			A26	Insufficient Device Problem Information		
5090	2017	All other reportable incidents			CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A15	Activation, Positioning or Separation Problem		
									A150103	Premature Activation		
5091	2017	TBD	35223	Endocardial/interventricular septal pacing lead	CT1684	Cardiac pacemakers and associated devices	F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A0104	Migration or Expulsion of Device		
							F1905	Device Revision or Replacement	A0104	Migration or Expulsion of Device		
5092	2018	All other reportable incidents	35223	Endocardial/interventricular septal pacing lead	CT1684	Cardiac pacemakers and associated devices	F1905	Device Revision or Replacement	A15	Activation, Positioning or Separation Problem	03/10/2018	
									A150203	Difficult or Delayed Positioning	03/10/2018	
5093	2017	TBD	35223	Endocardial/interventricular septal pacing lead	CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A15	Activation, Positioning or Separation Problem	24/10/2017	
									A150203	Difficult or Delayed Positioning	24/10/2017	
5094	2018	All other reportable incidents			CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A1104	Computer Operating System Problem		
5095	2017	TBD			CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A07	Electrical /Electronic Property Problem		
									A0709	Device Sensing Problem		
5096	2017	TBD			CT257	Implantable cardiac pacemakers	F1905	Device Revision or Replacement	A11	Computer Software Problem		
									A1102	Application Program Problem		
5097	2017	All other reportable incidents	47270	Cardiac resynchronization therapy implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A07	Electrical /Electronic Property Problem		
									A0709	Device Sensing Problem		
5098	2018	All other reportable incidents			CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A07	Electrical /Electronic Property Problem	01/09/2018	
									A0709	Device Sensing Problem	01/09/2018	
5099	2018	All other reportable incidents	35852	Single-chamber implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A07	Electrical /Electronic Property Problem		
									A0709	Device Sensing Problem		
5100	2018	All other reportable incidents	35852	Single-chamber implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A07	Electrical /Electronic Property Problem		
									A0709	Device Sensing Problem		
5101	2017	All other reportable incidents	35852	Single-chamber implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A07	Electrical /Electronic Property Problem		
									A0708	Power Problem		
5102	2017	TBD	35852	Single-chamber implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E0402	Hypersensitivity/Allergic Reaction	A01	Patient Device Interaction Problem		
									A17	Compatibility Problem		
5103	2018	All other reportable incidents			CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A07	Electrical /Electronic Property Problem		
									A0709	Device Sensing Problem		
5104	2018	All other reportable incidents			CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A1104	Computer Operating System Problem		
5105	2018	All other reportable incidents			CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A07	Electrical /Electronic Property Problem		
									A0709	Device Sensing Problem		
5106	2017	TBD	35852	Single-chamber implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	F05	Delay to Treatment/ Therapy	A07	Electrical /Electronic Property Problem		

									A0709	Device Sensing Problem		
							F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A07	Electrical /Electronic Property Problem		
									A0709	Device Sensing Problem		
5107	2018	All other reportable incidents	47270	Cardiac resynchronization therapy implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A1304	Telemetry Discrepancy		
									A16	Protective Measures Problem		
									A1602	Fail-Safe Problem		
									A26	Insufficient Device Problem Information		
5108	2017	All other reportable incidents			CT257	Implantable cardiac pacemakers	F1905	Device Revision or Replacement	A07	Electrical /Electronic Property Problem	01/03/2017	
									A0708	Power Problem	01/03/2017	
5109	2017	All other reportable incidents	47270	Cardiac resynchronization therapy implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A07	Electrical /Electronic Property Problem		
									A0709	Device Sensing Problem		
5110	2017	TBD	47270	Cardiac resynchronization therapy implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A07	Electrical /Electronic Property Problem		
									A0708	Power Problem		
5111	2018	Unanticipated serious deterioration in state of health			CT257	Implantable cardiac pacemakers			A26	Insufficient Device Problem Information		
5112	2018	All other reportable incidents	47270	Cardiac resynchronization therapy implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A09	Output Problem		
									A090404	Output Above Specifications		
5113	2018	All other reportable incidents	35852	Single-chamber implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	F1905	Device Revision or Replacement	A01	Patient Device Interaction Problem		
									A0104	Migration or Expulsion of Device		
									A17	Compatibility Problem		
5114	2018	All other reportable incidents			CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A07	Electrical /Electronic Property Problem	03/01/2018	
									A0708	Power Problem	03/01/2018	
5115	2017	TBD	35223	Endocardial/interventricular septal pacing lead	CT1684	Cardiac pacemakers and associated devices	F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A26	Insufficient Device Problem Information		
5116	2018	All other reportable incidents	47270	Cardiac resynchronization therapy implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	F1905	Device Revision or Replacement	A07	Electrical /Electronic Property Problem	13/12/2017	
									A0708	Power Problem	13/12/2017	
5117	2017	All other reportable incidents			CT257	Implantable cardiac pacemakers	F1905	Device Revision or Replacement	A07	Electrical /Electronic Property Problem	01/03/2017	
									A0708	Power Problem	01/03/2017	
5118	2018	TBD			CT257	Implantable cardiac pacemakers	F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A07	Electrical /Electronic Property Problem		
									A0709	Device Sensing Problem		
									A09	Output Problem		
									A0908	Incorrect, Inadequate or Imprecise Result or Readings		
							F1905	Device Revision or Replacement	A07	Electrical /Electronic Property Problem		
									A0709	Device Sensing Problem		
									A09	Output Problem		
									A0908	Incorrect, Inadequate or Imprecise Result or Readings		
5119	2018	All other reportable incidents	47270	Cardiac resynchronization therapy implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	F1905	Device Revision or Replacement	A07	Electrical /Electronic Property Problem		
									A0708	Power Problem		
5120	2017	All other reportable incidents			CT257	Implantable cardiac pacemakers	F1905	Device Revision or Replacement	A07	Electrical /Electronic Property Problem		
									A0708	Power Problem		
									A11	Computer Software Problem		
									A1102	Application Program Problem		
5121	2018	TBD	47270	Cardiac resynchronization therapy implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	F1905	Device Revision or Replacement	A07	Electrical /Electronic Property Problem		
									A0708	Power Problem		
5122	2018	All other reportable incidents	47270	Cardiac resynchronization therapy implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A07	Electrical /Electronic Property Problem	23/05/2018	
									A0708	Power Problem	23/05/2018	
5123	2018	All other reportable incidents	35223	Endocardial/interventricular septal pacing lead	CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A07	Electrical /Electronic Property Problem		
									A0709	Device Sensing Problem		
5124	2017	All other reportable incidents			CT257	Implantable cardiac pacemakers	F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A07	Electrical /Electronic Property Problem		
									A0709	Device Sensing Problem		
5125	2018	All other reportable incidents	47270	Cardiac resynchronization therapy implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	F1905	Device Revision or Replacement	A07	Electrical /Electronic Property Problem	23/05/2018	
									A0708	Power Problem	23/05/2018	
5126	2018	All other reportable incidents			CT257	Implantable cardiac pacemakers	F05	Delay to Treatment/ Therapy	A07	Electrical /Electronic Property Problem		
									A0709	Device Sensing Problem		
5127	2017	TBD	35223	Endocardial/interventricular septal pacing lead	CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A07	Electrical /Electronic Property Problem		
									A0709	Device Sensing Problem		
									A12	Connection Problem		
5128	2018	All other reportable incidents	35223	Endocardial/interventricular septal pacing lead	CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A05	Mechanical Problem	27/09/2018	
									A0510	Retraction Problem	27/09/2018	
5129	2018	All other reportable incidents	47270	Cardiac resynchronization therapy implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A07	Electrical /Electronic Property Problem	28/12/2017	
									A0708	Power Problem	28/12/2017	
5130	2017	All other reportable incidents	35223	Endocardial/interventricular septal pacing lead	CT1684	Cardiac pacemakers and associated devices	F1905	Device Revision or Replacement	A07	Electrical /Electronic Property Problem		
									A0709	Device Sensing Problem		
5131	2018	All other reportable incidents			CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A07	Electrical /Electronic Property Problem		
									A0708	Power Problem		
5132	2018	All other reportable incidents	35223	Endocardial/interventricular septal pacing lead	CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A05	Mechanical Problem	28/09/2018	
									A0510	Retraction Problem	28/09/2018	
5133	2018	All other reportable incidents	35223	Endocardial/interventricular septal pacing lead	CT1684	Cardiac pacemakers and associated devices	F1905	Device Revision or Replacement	A07	Electrical /Electronic Property Problem		
									A0709	Device Sensing Problem		
5134	2018	TBD	47270	Cardiac resynchronization therapy implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A07	Electrical /Electronic Property Problem		
									A0709	Device Sensing Problem		
5135	2018	TBD	47270	Cardiac resynchronization therapy implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A07	Electrical /Electronic Property Problem		
									A0708	Power Problem		
5136	2018	TBD	35852	Single-chamber implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A09	Output Problem		
									A090404	Output Above Specifications		
5137	2018	TBD			CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A07	Electrical /Electronic Property Problem		
									A0709	Device Sensing Problem		
5138	2018	All other reportable incidents	47270	Cardiac resynchronization therapy implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	F1905	Device Revision or Replacement	A07	Electrical /Electronic Property Problem		
									A0708	Power Problem		
5139	2018	All other reportable incidents	35852	Single-chamber implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A07	Electrical /Electronic Property Problem		
									A0709	Device Sensing Problem		
5140	2018	All other reportable incidents	35223	Endocardial/interventricular septal pacing lead	CT1684	Cardiac pacemakers and associated devices	F05	Delay to Treatment/ Therapy	A05	Mechanical Problem		
									A0510	Retraction Problem		
									A07	Electrical /Electronic Property Problem		
									A0709	Device Sensing Problem		
5141	2017	All other reportable incidents	47270	Cardiac resynchronization therapy implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A11	Computer Software Problem		
									A1102	Application Program Problem		

5142	2017	All other reportable incidents			CT257	Implantable cardiac pacemakers	F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A07	Electrical /Electronic Property Problem	01/03/2017	
									A0708	Power Problem	01/03/2017	
5143	2017	All other reportable incidents			CT257	Implantable cardiac pacemakers	F1905	Device Revision or Replacement	A07	Electrical /Electronic Property Problem	01/03/2017	
									A0708	Power Problem	01/03/2017	
5144	2018	All other reportable incidents	47270	Cardiac resynchronization therapy implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E2104	Electric Shock	A07	Electrical /Electronic Property Problem		
									A0709	Device Sensing Problem		
5145	2018	All other reportable incidents	47270	Cardiac resynchronization therapy implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	F1905	Device Revision or Replacement	A07	Electrical /Electronic Property Problem	09/06/2018	
									A0708	Power Problem	09/06/2018	
5146	2018	All other reportable incidents			CT257	Implantable cardiac pacemakers	F1905	Device Revision or Replacement	A07	Electrical /Electronic Property Problem		
									A0708	Power Problem		
5147	2017	TBD	47263	Cardiac resynchronization therapy implantable pacemaker	CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A07	Electrical /Electronic Property Problem		
									A0708	Power Problem		
					CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A07	Electrical /Electronic Property Problem		
									A0708	Power Problem		
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A07	Electrical /Electronic Property Problem		
									A0708	Power Problem		
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A07	Electrical /Electronic Property Problem		
									A0708	Power Problem		
5148	2018	Death			CT257	Implantable cardiac pacemakers	E2401	Insufficient Information	A07	Electrical /Electronic Property Problem	12/10/2018	
									A0708	Power Problem	12/10/2018	
5149	2018	All other reportable incidents	47270	Cardiac resynchronization therapy implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A07	Electrical /Electronic Property Problem	04/01/2018	
									A0708	Power Problem	04/01/2018	
5150	2018	All other reportable incidents			CT1684	Cardiac pacemakers and associated devices	F1905	Device Revision or Replacement	A07	Electrical /Electronic Property Problem		
									A0708	Power Problem		
5151	2017	All other reportable incidents	47270	Cardiac resynchronization therapy implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	F1905	Device Revision or Replacement	A07	Electrical /Electronic Property Problem	10/02/2017	
									A0708	Power Problem	10/02/2017	
5152	2018	All other reportable incidents	47270	Cardiac resynchronization therapy implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A1304	Telemetry Discrepancy		
									A26	Insufficient Device Problem Information		
5153	2017	TBD			CT257	Implantable cardiac pacemakers	F1905	Device Revision or Replacement	A07	Electrical /Electronic Property Problem		
									A0708	Power Problem		
5154	2017	All other reportable incidents			CT257	Implantable cardiac pacemakers	F1905	Device Revision or Replacement	A07	Electrical /Electronic Property Problem	01/03/2017	
									A0708	Power Problem	01/03/2017	
5155	2018	All other reportable incidents			CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A09	Output Problem		
									A0907	No Device Output		
5156	2018	All other reportable incidents	47270	Cardiac resynchronization therapy implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A07	Electrical /Electronic Property Problem		
									A0709	Device Sensing Problem		
5157	2018	All other reportable incidents			CT257	Implantable cardiac pacemakers	F1905	Device Revision or Replacement	A13	Communication or Transmission Problem	12/11/2017	
									A26	Insufficient Device Problem Information	12/11/2017	
5158	2018	All other reportable incidents	47270	Cardiac resynchronization therapy implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A07	Electrical /Electronic Property Problem		
									A0709	Device Sensing Problem		
5159	2018	TBD	35223	Endocardial/interventricular septal pacing lead	CT1684	Cardiac pacemakers and associated devices			A26	Insufficient Device Problem Information		
5160	2017	All other reportable incidents	47270	Cardiac resynchronization therapy implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A07	Electrical /Electronic Property Problem		
									A0709	Device Sensing Problem		
5161	2017	All other reportable incidents			CT257	Implantable cardiac pacemakers	F1905	Device Revision or Replacement	A07	Electrical /Electronic Property Problem	23/02/2017	
									A0708	Power Problem	23/02/2017	
									A13	Communication or Transmission Problem	23/02/2017	
									A26	Insufficient Device Problem Information	23/02/2017	
5162	2017	TBD	35223	Endocardial/interventricular septal pacing lead	CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A15	Activation, Positioning or Separation Problem		
									A150101	Activation Failure		
5163	2018	All other reportable incidents	47270	Cardiac resynchronization therapy implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A07	Electrical /Electronic Property Problem		
									A0709	Device Sensing Problem		
5164	2018	All other reportable incidents	47270	Cardiac resynchronization therapy implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A07	Electrical /Electronic Property Problem	23/05/2018	
									A0708	Power Problem	23/05/2018	
5165	2018	TBD	35223	Endocardial/interventricular septal pacing lead	CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A15	Activation, Positioning or Separation Problem		
									A150101	Activation Failure		
									A150203	Difficult or Delayed Positioning		
5166	2018	TBD	35223	Endocardial/interventricular septal pacing lead	CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A15	Activation, Positioning or Separation Problem		
									A150203	Difficult or Delayed Positioning		
5167	2018	TBD	35852	Single-chamber implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A07	Electrical /Electronic Property Problem		
									A0709	Device Sensing Problem		
5168	2017	TBD	35223	Endocardial/interventricular septal pacing lead	CT1684	Cardiac pacemakers and associated devices	E0402	Hypersensitivity/Allergic Reaction	A26	Insufficient Device Problem Information		
5169	2017	All other reportable incidents			CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A07	Electrical /Electronic Property Problem		
									A0708	Power Problem		
5170	2017	TBD	47270	Cardiac resynchronization therapy implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A07	Electrical /Electronic Property Problem		
									A0709	Device Sensing Problem		
5171	2017	TBD	35852	Single-chamber implantable defibrillator	CT2866	Implantable cardiac electrical stimulators			A07	Electrical /Electronic Property Problem		
									A0709	Device Sensing Problem		
5172	2017	All other reportable incidents	35223	Endocardial/interventricular septal pacing lead	CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A05	Mechanical Problem		
									A0506	Mechanical Jam		
5173	2018	All other reportable incidents	35852	Single-chamber implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	F1905	Device Revision or Replacement	A01	Patient Device Interaction Problem	01/10/2018	
									A17	Compatibility Problem	01/10/2018	
5174	2018	All other reportable incidents	47270	Cardiac resynchronization therapy implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A07	Electrical /Electronic Property Problem		
									A0708	Power Problem		
5175	2017	TBD	47270	Cardiac resynchronization therapy implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	F1905	Device Revision or Replacement	A07	Electrical /Electronic Property Problem		
									A0708	Power Problem		
5176	2018	All other reportable incidents	35852	Single-chamber implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A07	Electrical /Electronic Property Problem		
									A0713	Defibrillation/Stimulation Problem		
									A071302	Inappropriate/Inadequate Shock/Stimulation		
5177	2018	All other reportable incidents			CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A07	Electrical /Electronic Property Problem		
									A0708	Power Problem		
5178	2018	All other reportable incidents	35852	Single-chamber implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A07	Electrical /Electronic Property Problem		
									A0709	Device Sensing Problem		
									A070909	Oversensing		
5179	2018	All other reportable incidents	35852	Single-chamber implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A07	Electrical /Electronic Property Problem		
									A0709	Device Sensing Problem		
5180	2017	TBD			CT257	Implantable cardiac pacemakers	E0402	Hypersensitivity/Allergic Reaction	A01	Patient Device Interaction Problem		

									A17	Compatibility Problem		
							E2403	No Clinical Signs, Symptoms or Conditions	A01	Patient Device Interaction Problem		
									A17	Compatibility Problem		
5181	2017	All other reportable incidents			CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A07	Electrical /Electronic Property Problem	26/07/2017	
									A0709	Device Sensing Problem	26/07/2017	
5182	2018	All other reportable incidents	47270	Cardiac resynchronization therapy implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	F1905	Device Revision or Replacement	A07	Electrical /Electronic Property Problem	22/12/2017	
									A0708	Power Problem	22/12/2017	
5183	2018	All other reportable incidents			CT257	Implantable cardiac pacemakers	F1905	Device Revision or Replacement	A04	Material Integrity Problem	21/09/2018	
									A0413	Material Separation	21/09/2018	
5184	2017	TBD	35223	Endocardial/interventricular septal pacing lead	CT1684	Cardiac pacemakers and associated devices			A26	Insufficient Device Problem Information		
5185	2018	TBD	47270	Cardiac resynchronization therapy implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A07	Electrical /Electronic Property Problem		
									A0709	Device Sensing Problem		
5186	2017	TBD	47270	Cardiac resynchronization therapy implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A07	Electrical /Electronic Property Problem		
									A0708	Power Problem		
5187	2018	All other reportable incidents	35223	Endocardial/interventricular septal pacing lead	CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A0104	Migration or Expulsion of Device	18/12/2017	
									A15	Activation, Positioning or Separation Problem	18/12/2017	
									A150101	Activation Failure	18/12/2017	
5188	2017	TBD			CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A07	Electrical /Electronic Property Problem		
									A0709	Device Sensing Problem		
5189	2018	All other reportable incidents	35852	Single-chamber implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E2104	Electric Shock	A07	Electrical /Electronic Property Problem		
									A0709	Device Sensing Problem		
5190	2017	All other reportable incidents	35223	Endocardial/interventricular septal pacing lead	CT1684	Cardiac pacemakers and associated devices	F14	Prolonged Episode of Care	A07	Electrical /Electronic Property Problem		
									A0709	Device Sensing Problem		
5191	2017	All other reportable incidents	47270	Cardiac resynchronization therapy implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A09	Output Problem		
									A0907	No Device Output		
5192	2017	All other reportable incidents	47270	Cardiac resynchronization therapy implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	F1905	Device Revision or Replacement	A07	Electrical /Electronic Property Problem	07/03/2017	
									A0708	Power Problem	07/03/2017	
5193	2017	All other reportable incidents			CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A07	Electrical /Electronic Property Problem	28/02/2017	
									A0708	Power Problem	28/02/2017	
5194	2018	All other reportable incidents			CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A07	Electrical /Electronic Property Problem	14/10/2018	
									A0708	Power Problem	14/10/2018	
5195	2018	All other reportable incidents	47270	Cardiac resynchronization therapy implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A07	Electrical /Electronic Property Problem	20/12/2017	
									A0708	Power Problem	20/12/2017	
5196	2018	All other reportable incidents			CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A05	Mechanical Problem		
									A0506	Mechanical Jam		
5197	2018	All other reportable incidents	47270	Cardiac resynchronization therapy implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	F1905	Device Revision or Replacement	A07	Electrical /Electronic Property Problem	24/10/2018	
									A0708	Power Problem	24/10/2018	
5198	2018	All other reportable incidents	47270	Cardiac resynchronization therapy implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A07	Electrical /Electronic Property Problem		
									A0709	Device Sensing Problem		
5199	2018	All other reportable incidents	35223	Endocardial/interventricular septal pacing lead	CT1684	Cardiac pacemakers and associated devices	F14	Prolonged Episode of Care	A0104	Migration or Expulsion of Device	25/09/2018	
5200	2018	All other reportable incidents	35223	Endocardial/interventricular septal pacing lead	CT1684	Cardiac pacemakers and associated devices	F1905	Device Revision or Replacement	A05	Mechanical Problem	04/10/2018	
									A0510	Retraction Problem	04/10/2018	
5201	2017	TBD	35852	Single-chamber implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	F1905	Device Revision or Replacement	A07	Electrical /Electronic Property Problem		
									A0709	Device Sensing Problem		
									A09	Output Problem		
									A090404	Output Above Specifications		
5202	2017	All other reportable incidents	47270	Cardiac resynchronization therapy implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A07	Electrical /Electronic Property Problem		
									A0708	Power Problem		
5203	2017	TBD	47270	Cardiac resynchronization therapy implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A07	Electrical /Electronic Property Problem		
									A0708	Power Problem		
5204	2018	All other reportable incidents	47270	Cardiac resynchronization therapy implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	F1905	Device Revision or Replacement	A07	Electrical /Electronic Property Problem	31/05/2018	
									A0708	Power Problem	31/05/2018	
5205	2018	All other reportable incidents	35852	Single-chamber implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A07	Electrical /Electronic Property Problem		
									A0709	Device Sensing Problem		
5206	2018	All other reportable incidents	47270	Cardiac resynchronization therapy implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A07	Electrical /Electronic Property Problem		
									A0703	Arcing		
5207	2018	All other reportable incidents	47270	Cardiac resynchronization therapy implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E2401	Insufficient Information	A07	Electrical /Electronic Property Problem		
									A0708	Power Problem		
5208	2017	TBD	47270	Cardiac resynchronization therapy implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	F1905	Device Revision or Replacement	A07	Electrical /Electronic Property Problem		
									A0708	Power Problem		
5209	2018	All other reportable incidents	47270	Cardiac resynchronization therapy implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	F1905	Device Revision or Replacement	A07	Electrical /Electronic Property Problem		
									A0709	Device Sensing Problem		
5210	2018	All other reportable incidents	47270	Cardiac resynchronization therapy implantable defibrillator	CT2866	Implantable cardiac electrical stimulators			A07	Electrical /Electronic Property Problem		
									A0709	Device Sensing Problem		
5211	2017	TBD			CT257	Implantable cardiac pacemakers	F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A0104	Migration or Expulsion of Device		
									A07	Electrical /Electronic Property Problem		
									A0709	Device Sensing Problem		
									A23	Use of Device Problem		
							F1905	Device Revision or Replacement	A0104	Migration or Expulsion of Device		
									A07	Electrical /Electronic Property Problem		
									A0709	Device Sensing Problem		
									A23	Use of Device Problem		
5212	2017	TBD			CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A07	Electrical /Electronic Property Problem	03/08/2017	
									A0709	Device Sensing Problem	03/08/2017	
									A09	Output Problem	03/08/2017	
									A090404	Output Above Specifications	03/08/2017	
									A090502	Radiation Overexposure	03/08/2017	
5213	2018	All other reportable incidents			CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A13	Communication or Transmission Problem		
									A26	Insufficient Device Problem Information		
5214	2018	All other reportable incidents	47270	Cardiac resynchronization therapy implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A07	Electrical /Electronic Property Problem		
									A0709	Device Sensing Problem		
5215	2018	All other reportable incidents	35852	Single-chamber implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A07	Electrical /Electronic Property Problem		
									A0709	Device Sensing Problem		
5216	2018	All other reportable incidents			CT257	Implantable cardiac pacemakers	F1905	Device Revision or Replacement	A07	Electrical /Electronic Property Problem	26/10/2018	
									A0708	Power Problem	26/10/2018	
5217	2017	All other reportable incidents			CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A13	Communication or Transmission Problem		

									A26	Insufficient Device Problem Information			
5218	2018	All other reportable incidents	35223	Endocardial/interventricular septal pacing lead	CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A05	Mechanical Problem			
									A0510	Retraction Problem			
5219	2018	Unanticipated serious deterioration in state of health	35223	Endocardial/interventricular septal pacing lead	CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A05	Mechanical Problem			
									A0510	Retraction Problem			
5220	2018	All other reportable incidents	47270	Cardiac resynchronization therapy implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	F1905	Device Revision or Replacement	A07	Electrical /Electronic Property Problem	15/06/2018		
									A0708	Power Problem	15/06/2018		
5221	2017	All other reportable incidents			CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A07	Electrical /Electronic Property Problem			
									A0708	Power Problem			
5222	2017	TBD			CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A07	Electrical /Electronic Property Problem			
									A0709	Device Sensing Problem			
5223	2017	All other reportable incidents	35223	Endocardial/interventricular septal pacing lead	CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A15	Activation, Positioning or Separation Problem			
									A150203	Difficult or Delayed Positioning			
5224	2017	TBD	35223	Endocardial/interventricular septal pacing lead	CT1684	Cardiac pacemakers and associated devices	E0402	Hypersensitivity/Allergic Reaction	A26	Insufficient Device Problem Information			
5225	2018	All other reportable incidents			CT257	Implantable cardiac pacemakers	F1905	Device Revision or Replacement	A07	Electrical /Electronic Property Problem			
									A0709	Device Sensing Problem			
									A09	Output Problem			
									A0907	No Device Output			
5226	2018	All other reportable incidents	35223	Endocardial/interventricular septal pacing lead	CT1684	Cardiac pacemakers and associated devices	E2401	Insufficient Information	A26	Insufficient Device Problem Information			
5227	2017	TBD			CT257	Implantable cardiac pacemakers	F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A05	Mechanical Problem			
									A0506	Mechanical Jam			
5228	2018	All other reportable incidents	47270	Cardiac resynchronization therapy implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A11	Computer Software Problem			
									A1102	Application Program Problem			
5229	2017	All other reportable incidents			CT257	Implantable cardiac pacemakers	F1905	Device Revision or Replacement	A07	Electrical /Electronic Property Problem	13/03/2017		
									A0708	Power Problem	13/03/2017		
5230	2017	All other reportable incidents			CT257	Implantable cardiac pacemakers	F14	Prolonged Episode of Care	A12	Connection Problem	22/02/2017		
									A1204	Failure to Disconnect	22/02/2017		
							F1905	Device Revision or Replacement	A12	Connection Problem	22/02/2017		
									A1204	Failure to Disconnect	22/02/2017		
5231	2017	All other reportable incidents	47270	Cardiac resynchronization therapy implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	F1905	Device Revision or Replacement	A07	Electrical /Electronic Property Problem	20/02/2017		
									A0708	Power Problem	20/02/2017		
5232	2018	All other reportable incidents	47270	Cardiac resynchronization therapy implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A07	Electrical /Electronic Property Problem			
									A0708	Power Problem			
5233	2018	All other reportable incidents	47270	Cardiac resynchronization therapy implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A07	Electrical /Electronic Property Problem	25/10/2018		
									A0708	Power Problem	25/10/2018		
5234	2018	All other reportable incidents	47270	Cardiac resynchronization therapy implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A07	Electrical /Electronic Property Problem			
									A0708	Power Problem			
									A13	Communication or Transmission Problem			
									A26	Insufficient Device Problem Information			
5235	2018	All other reportable incidents	47270	Cardiac resynchronization therapy implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	F1905	Device Revision or Replacement	A07	Electrical /Electronic Property Problem	23/10/2018		
									A0708	Power Problem	23/10/2018		
5236	2017	All other reportable incidents			CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A07	Electrical /Electronic Property Problem			
									A0708	Power Problem			
5237	2018	All other reportable incidents	47270	Cardiac resynchronization therapy implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	F1905	Device Revision or Replacement	A07	Electrical /Electronic Property Problem	21/05/2018		
									A0708	Power Problem	21/05/2018		
5238	2017	TBD			CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A07	Electrical /Electronic Property Problem			
									A0708	Power Problem			
5239	2017	All other reportable incidents			CT257	Implantable cardiac pacemakers	F1905	Device Revision or Replacement	A07	Electrical /Electronic Property Problem			
									A0709	Device Sensing Problem			
5240	2018	TBD			CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A07	Electrical /Electronic Property Problem			
									A0709	Device Sensing Problem			
5241	2017	All other reportable incidents	35223	Endocardial/interventricular septal pacing lead	CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A15	Activation, Positioning or Separation Problem			
									A150203	Difficult or Delayed Positioning			
5242	2018	All other reportable incidents	47270	Cardiac resynchronization therapy implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A07	Electrical /Electronic Property Problem			
									A0708	Power Problem			
5243	2017	Unanticipated serious deterioration in state of health	35223	Endocardial/interventricular septal pacing lead	CT1684	Cardiac pacemakers and associated devices	F14	Prolonged Episode of Care	A15	Activation, Positioning or Separation Problem			
									A150203	Difficult or Delayed Positioning			
5244	2017	TBD	35223	Endocardial/interventricular septal pacing lead	CT1684	Cardiac pacemakers and associated devices	E2401	Insufficient Information	A26	Insufficient Device Problem Information			
5245	2018	All other reportable incidents	35223	Endocardial/interventricular septal pacing lead	CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A07	Electrical /Electronic Property Problem			
									A0709	Device Sensing Problem			
5246	2018	Unanticipated serious deterioration in state of health	35223	Endocardial/interventricular septal pacing lead	CT1684	Cardiac pacemakers and associated devices	E2114	Perforation	A15	Activation, Positioning or Separation Problem	10/10/2018		
									A150203	Difficult or Delayed Positioning	10/10/2018		
							F1905	Device Revision or Replacement	A15	Activation, Positioning or Separation Problem	10/10/2018		
									A150203	Difficult or Delayed Positioning	10/10/2018		
5247	2018	All other reportable incidents			CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A07	Electrical /Electronic Property Problem			
									A0709	Device Sensing Problem			
5248	2017	All other reportable incidents			CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A09	Output Problem			
									A0907	No Device Output			
5249	2017	Death			CT1684	Cardiac pacemakers and associated devices			A07	Electrical /Electronic Property Problem			
									A0709	Device Sensing Problem			
									A09	Output Problem			
									A090404	Output Above Specifications			
									A1104	Computer Operating System Problem			
5250	2017	All other reportable incidents	35852	Single-chamber implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A07	Electrical /Electronic Property Problem			
									A0709	Device Sensing Problem			
5251	2018	All other reportable incidents	47270	Cardiac resynchronization therapy implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A07	Electrical /Electronic Property Problem			
									A0709	Device Sensing Problem			
5252	2018	Unanticipated serious deterioration in state of health	35223	Endocardial/interventricular septal pacing lead	CT1684	Cardiac pacemakers and associated devices			A26	Insufficient Device Problem Information			
5253	2017	TBD	35223	Endocardial/interventricular septal pacing lead	CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A07	Electrical /Electronic Property Problem			
									A0709	Device Sensing Problem			
5254	2018	All other reportable incidents			CT257	Implantable cardiac pacemakers	F1905	Device Revision or Replacement	A07	Electrical /Electronic Property Problem			
									A0708	Power Problem			
5255	2017	TBD			CT257	Implantable cardiac pacemakers			A26	Insufficient Device Problem Information			
5256	2017	TBD			CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A07	Electrical /Electronic Property Problem			
									A0709	Device Sensing Problem			

5257	2017	TBD	35852	Single-chamber implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E1906	Unspecified Infection	A01	Patient Device Interaction Problem		
									A17	Compatibility Problem		
5258	2017	All other reportable incidents	47270	Cardiac resynchronization therapy implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	F1905	Device Revision or Replacement	A07	Electrical /Electronic Property Problem		
									A0708	Power Problem		
5259	2017	TBD	35852	Single-chamber implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A07	Electrical /Electronic Property Problem		
									A0709	Device Sensing Problem		
							F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A07	Electrical /Electronic Property Problem		
									A0709	Device Sensing Problem		
5260	2017	TBD	35223	Endocardial/interventricular septal pacing lead	CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A04	Material Integrity Problem		
									A0404	Crack		
5261	2018	All other reportable incidents	47270	Cardiac resynchronization therapy implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A11	Computer Software Problem		
									A1102	Application Program Problem		
									A1304	Telemetry Discrepancy		
									A26	Insufficient Device Problem Information		
5262	2017	All other reportable incidents	47263	Cardiac resynchronization therapy implantable pacemaker	CT1684	Cardiac pacemakers and associated devices	F1905	Device Revision or Replacement	A09	Output Problem	12/04/2017	
									A0908	Incorrect, Inadequate or Imprecise Result or Readings	12/04/2017	
					CT1685	Cardiac pacemakers	F1905	Device Revision or Replacement	A09	Output Problem	12/04/2017	
									A0908	Incorrect, Inadequate or Imprecise Result or Readings	12/04/2017	
					CT257	Implantable cardiac pacemakers	F1905	Device Revision or Replacement	A09	Output Problem	12/04/2017	
									A0908	Incorrect, Inadequate or Imprecise Result or Readings	12/04/2017	
					CT2866	Implantable cardiac electrical stimulators	F1905	Device Revision or Replacement	A09	Output Problem	12/04/2017	
									A0908	Incorrect, Inadequate or Imprecise Result or Readings	12/04/2017	
5263	2018	All other reportable incidents	35223	Endocardial/interventricular septal pacing lead	CT1684	Cardiac pacemakers and associated devices	E2401	Insufficient Information	A26	Insufficient Device Problem Information		
5264	2018	All other reportable incidents	47270	Cardiac resynchronization therapy implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A09	Output Problem		
									A090404	Output Above Specifications		
5265	2018	All other reportable incidents	47263	Cardiac resynchronization therapy implantable pacemaker	CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A07	Electrical /Electronic Property Problem		
									A0709	Device Sensing Problem		
					CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A07	Electrical /Electronic Property Problem		
									A0709	Device Sensing Problem		
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A07	Electrical /Electronic Property Problem		
									A0709	Device Sensing Problem		
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A07	Electrical /Electronic Property Problem		
									A0709	Device Sensing Problem		
5266	2017	All other reportable incidents			CT257	Implantable cardiac pacemakers	F1905	Device Revision or Replacement	A07	Electrical /Electronic Property Problem	03/03/2017	
									A0708	Power Problem	03/03/2017	
5267	2018	All other reportable incidents	35223	Endocardial/interventricular septal pacing lead	CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A07	Electrical /Electronic Property Problem	08/06/2018	
									A0709	Device Sensing Problem	08/06/2018	
5268	2018	All other reportable incidents	35223	Endocardial/interventricular septal pacing lead	CT1684	Cardiac pacemakers and associated devices	F1905	Device Revision or Replacement	A07	Electrical /Electronic Property Problem		
									A0709	Device Sensing Problem		
5269	2018	All other reportable incidents	47270	Cardiac resynchronization therapy implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A11	Computer Software Problem		
									A1102	Application Program Problem		
5270	2017	All other reportable incidents			CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A05	Mechanical Problem		
									A0504	Leak/Splash		
									A07	Electrical /Electronic Property Problem		
									A0721	Circuit Failure		
5271	2018	TBD	35852	Single-chamber implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A07	Electrical /Electronic Property Problem		
									A0709	Device Sensing Problem		
5272	2017	TBD	47270	Cardiac resynchronization therapy implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A07	Electrical /Electronic Property Problem		
									A0709	Device Sensing Problem		
5273	2018	All other reportable incidents	35223	Endocardial/interventricular septal pacing lead	CT1684	Cardiac pacemakers and associated devices	F1905	Device Revision or Replacement	A01	Patient Device Interaction Problem		
									A17	Compatibility Problem		
5274	2017	TBD	47270	Cardiac resynchronization therapy implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A07	Electrical /Electronic Property Problem		
									A0708	Power Problem		
5275	2017	TBD	35223	Endocardial/interventricular septal pacing lead	CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A0104	Migration or Expulsion of Device		
5276	2017	TBD	35223	Endocardial/interventricular septal pacing lead	CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A15	Activation, Positioning or Separation Problem		
									A150101	Activation Failure		
5277	2017	TBD	35223	Endocardial/interventricular septal pacing lead	CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A15	Activation, Positioning or Separation Problem		
									A150101	Activation Failure		
5278	2017	TBD	35852	Single-chamber implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A07	Electrical /Electronic Property Problem		
									A0709	Device Sensing Problem		
5279	2017	All other reportable incidents	47270	Cardiac resynchronization therapy implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A26	Insufficient Device Problem Information		
5280	2018	All other reportable incidents	47270	Cardiac resynchronization therapy implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E0402	Hypersensitivity/Allergic Reaction	A11	Computer Software Problem		
									A1102	Application Program Problem		
							F05	Delay to Treatment/ Therapy	A11	Computer Software Problem		
									A1102	Application Program Problem		
5281	2018	All other reportable incidents	47270	Cardiac resynchronization therapy implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A09	Output Problem		
									A0908	Incorrect, Inadequate or Imprecise Result or Readings		
5282	2018	All other reportable incidents	60789	Intracardiac pacemaker	CT130	Catheters and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A07	Electrical /Electronic Property Problem		
									A0708	Power Problem		
					CT145	Cardiovascular devices	E2403	No Clinical Signs, Symptoms or Conditions	A07	Electrical /Electronic Property Problem		
									A0708	Power Problem		
					CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A07	Electrical /Electronic Property Problem		
									A0708	Power Problem		
					CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A07	Electrical /Electronic Property Problem		
									A0708	Power Problem		
					CT233	Surgical	E2403	No Clinical Signs, Symptoms or Conditions	A07	Electrical /Electronic Property Problem		
									A0708	Power Problem		
					CT2406	Implantable	E2403	No Clinical Signs, Symptoms or Conditions	A07	Electrical /Electronic Property Problem		
									A0708	Power Problem		
					CT2407	Active implantable	E2403	No Clinical Signs, Symptoms or Conditions	A07	Electrical /Electronic Property Problem		
									A0708	Power Problem		
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A07	Electrical /Electronic Property Problem		

									A0708	Power Problem		
					CT2727	Active implant devices/ancillaries	E2403	No Clinical Signs, Symptoms or Conditions	A07	Electrical /Electronic Property Problem		
									A0708	Power Problem		
					CT2750	Implantable cardiac electrical stimulators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A07	Electrical /Electronic Property Problem		
									A0708	Power Problem		
					CT2818	Catheterization/Cannulation/Fluid path devices	E2403	No Clinical Signs, Symptoms or Conditions	A07	Electrical /Electronic Property Problem		
									A0708	Power Problem		
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A07	Electrical /Electronic Property Problem		
									A0708	Power Problem		
					CT2985	Active	E2403	No Clinical Signs, Symptoms or Conditions	A07	Electrical /Electronic Property Problem		
									A0708	Power Problem		
					CT3001	Implantable metal	E2403	No Clinical Signs, Symptoms or Conditions	A07	Electrical /Electronic Property Problem		
									A0708	Power Problem		
					CT3002	Implantable synthetic polymer	E2403	No Clinical Signs, Symptoms or Conditions	A07	Electrical /Electronic Property Problem		
									A0708	Power Problem		
					CT3006	Implantable inorganic material	E2403	No Clinical Signs, Symptoms or Conditions	A07	Electrical /Electronic Property Problem		
									A0708	Power Problem		
					CT314	Electrically powered	E2403	No Clinical Signs, Symptoms or Conditions	A07	Electrical /Electronic Property Problem		
									A0708	Power Problem		
					CT321	Long-term surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A07	Electrical /Electronic Property Problem		
									A0708	Power Problem		
					CT981	Single use	E2403	No Clinical Signs, Symptoms or Conditions	A07	Electrical /Electronic Property Problem		
									A0708	Power Problem		
					CT983	Surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A07	Electrical /Electronic Property Problem		
									A0708	Power Problem		
5283	2018	All other reportable incidents			CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A11	Computer Software Problem		
									A1102	Application Program Problem		
5284	2018	All other reportable incidents	35223	Endocardial/interventricular septal pacing lead	CT1684	Cardiac pacemakers and associated devices	F1905	Device Revision or Replacement	A05	Mechanical Problem	25/07/2018	
									A0510	Retraction Problem	25/07/2018	
5285	2018	All other reportable incidents			CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A07	Electrical /Electronic Property Problem		
									A0708	Power Problem		
5286	2018	All other reportable incidents	60789	Intracardiac pacemaker	CT130	Catheters and associated devices	F14	Prolonged Episode of Care	A26	Insufficient Device Problem Information		
							F15	Recognized Device or Procedural Complication	A26	Insufficient Device Problem Information		
					CT145	Cardiovascular devices	F14	Prolonged Episode of Care	A26	Insufficient Device Problem Information		
							F15	Recognized Device or Procedural Complication	A26	Insufficient Device Problem Information		
					CT1684	Cardiac pacemakers and associated devices	F14	Prolonged Episode of Care	A26	Insufficient Device Problem Information		
							F15	Recognized Device or Procedural Complication	A26	Insufficient Device Problem Information		
					CT1685	Cardiac pacemakers	F14	Prolonged Episode of Care	A26	Insufficient Device Problem Information		
							F15	Recognized Device or Procedural Complication	A26	Insufficient Device Problem Information		
					CT233	Surgical	F14	Prolonged Episode of Care	A26	Insufficient Device Problem Information		
							F15	Recognized Device or Procedural Complication	A26	Insufficient Device Problem Information		
					CT2406	Implantable	F14	Prolonged Episode of Care	A26	Insufficient Device Problem Information		
							F15	Recognized Device or Procedural Complication	A26	Insufficient Device Problem Information		
					CT2407	Active implantable	F14	Prolonged Episode of Care	A26	Insufficient Device Problem Information		
							F15	Recognized Device or Procedural Complication	A26	Insufficient Device Problem Information		
					CT257	Implantable cardiac pacemakers	F14	Prolonged Episode of Care	A26	Insufficient Device Problem Information		
							F15	Recognized Device or Procedural Complication	A26	Insufficient Device Problem Information		
					CT2727	Active implant devices/ancillaries	F14	Prolonged Episode of Care	A26	Insufficient Device Problem Information		
							F15	Recognized Device or Procedural Complication	A26	Insufficient Device Problem Information		
					CT2750	Implantable cardiac electrical stimulators and associated devices	F14	Prolonged Episode of Care	A26	Insufficient Device Problem Information		
							F15	Recognized Device or Procedural Complication	A26	Insufficient Device Problem Information		
					CT2818	Catheterization/Cannulation/Fluid path devices	F14	Prolonged Episode of Care	A26	Insufficient Device Problem Information		
							F15	Recognized Device or Procedural Complication	A26	Insufficient Device Problem Information		
					CT2866	Implantable cardiac electrical stimulators	F14	Prolonged Episode of Care	A26	Insufficient Device Problem Information		
							F15	Recognized Device or Procedural Complication	A26	Insufficient Device Problem Information		
					CT2985	Active	F14	Prolonged Episode of Care	A26	Insufficient Device Problem Information		
							F15	Recognized Device or Procedural Complication	A26	Insufficient Device Problem Information		
					CT3001	Implantable metal	F14	Prolonged Episode of Care	A26	Insufficient Device Problem Information		
							F15	Recognized Device or Procedural Complication	A26	Insufficient Device Problem Information		
					CT3002	Implantable synthetic polymer	F14	Prolonged Episode of Care	A26	Insufficient Device Problem Information		
							F15	Recognized Device or Procedural Complication	A26	Insufficient Device Problem Information		
					CT3006	Implantable inorganic material	F14	Prolonged Episode of Care	A26	Insufficient Device Problem Information		
							F15	Recognized Device or Procedural Complication	A26	Insufficient Device Problem Information		
					CT314	Electrically powered	F14	Prolonged Episode of Care	A26	Insufficient Device Problem Information		
							F15	Recognized Device or Procedural Complication	A26	Insufficient Device Problem Information		
					CT321	Long-term surgical invasive	F14	Prolonged Episode of Care	A26	Insufficient Device Problem Information		
							F15	Recognized Device or Procedural Complication	A26	Insufficient Device Problem Information		
					CT981	Single use	F14	Prolonged Episode of Care	A26	Insufficient Device Problem Information		
							F15	Recognized Device or Procedural Complication	A26	Insufficient Device Problem Information		
					CT983	Surgical invasive	F14	Prolonged Episode of Care	A26	Insufficient Device Problem Information		
							F15	Recognized Device or Procedural Complication	A26	Insufficient Device Problem Information		
5287	2018	All other reportable incidents			CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A07	Electrical /Electronic Property Problem	08/05/2018	
									A0708	Power Problem	08/05/2018	
5288	2017	All other reportable incidents	35223	Endocardial/interventricular septal pacing lead	CT1684	Cardiac pacemakers and associated devices	F1905	Device Revision or Replacement	A04	Material Integrity Problem		
									A0405	Degraded		
									A07	Electrical /Electronic Property Problem		
									A0709	Device Sensing Problem		
5289	2018	All other reportable incidents			CT257	Implantable cardiac pacemakers	F1905	Device Revision or Replacement	A07	Electrical /Electronic Property Problem	05/09/2017	
									A0708	Power Problem	05/09/2017	
5290	2017	TBD	35223	Endocardial/interventricular septal pacing lead	CT1684	Cardiac pacemakers and associated devices	E2401	Insufficient Information	A26	Insufficient Device Problem Information		
5291	2018	All other reportable incidents			CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A09	Output Problem		
									A0907	No Device Output		
5292	2018	TBD			CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A07	Electrical /Electronic Property Problem		
									A0708	Power Problem		
5293	2018	TBD			CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A07	Electrical /Electronic Property Problem		
									A0708	Power Problem		
5294	2018	All other reportable incidents			CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A07	Electrical /Electronic Property Problem		

									A0708	Power Problem		
5295	2018	All other reportable incidents	35852	Single-chamber implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A07	Electrical /Electronic Property Problem		
									A0709	Device Sensing Problem		
5296	2018	All other reportable incidents	35223	Endocardial/interventricular septal pacing lead	CT1684	Cardiac pacemakers and associated devices	E2401	Insufficient Information	A26	Insufficient Device Problem Information		
5297	2017	TBD	35223	Endocardial/interventricular septal pacing lead	CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A0104	Migration or Expulsion of Device		
									A15	Activation, Positioning or Separation Problem		
									A150101	Activation Failure		
5298	2017	All other reportable incidents	47270	Cardiac resynchronization therapy implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	F1905	Device Revision or Replacement	A07	Electrical /Electronic Property Problem	03/03/2017	
									A0708	Power Problem	03/03/2017	
5299	2018	All other reportable incidents			CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A09	Output Problem		
									A0907	No Device Output		
5300	2018	All other reportable incidents			CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A09	Output Problem		
									A0907	No Device Output		
5301	2017	All other reportable incidents	47270	Cardiac resynchronization therapy implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A07	Electrical /Electronic Property Problem	07/03/2017	
									A0708	Power Problem	07/03/2017	
5302	2017	TBD	47270	Cardiac resynchronization therapy implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	F1905	Device Revision or Replacement	A07	Electrical /Electronic Property Problem		
									A0708	Power Problem		
5303	2018	All other reportable incidents			CT257	Implantable cardiac pacemakers	E0402	Hypersensitivity/Allergic Reaction	A07	Electrical /Electronic Property Problem		
									A0709	Device Sensing Problem		
5304	2018	TBD	35223	Endocardial/interventricular septal pacing lead	CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A07	Electrical /Electronic Property Problem		
									A0709	Device Sensing Problem		
5305	2018	All other reportable incidents	47270	Cardiac resynchronization therapy implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A07	Electrical /Electronic Property Problem	10/01/2018	
									A0709	Device Sensing Problem	10/01/2018	
5306	2017	TBD			CT257	Implantable cardiac pacemakers	F1905	Device Revision or Replacement	A07	Electrical /Electronic Property Problem		
									A0708	Power Problem		
5307	2017	TBD	47263	Cardiac resynchronization therapy implantable pacemaker	CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A07	Electrical /Electronic Property Problem		
									A0709	Device Sensing Problem		
					CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A07	Electrical /Electronic Property Problem		
									A0709	Device Sensing Problem		
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A07	Electrical /Electronic Property Problem		
									A0709	Device Sensing Problem		
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A07	Electrical /Electronic Property Problem		
									A0709	Device Sensing Problem		
5308	2018	All other reportable incidents	35852	Single-chamber implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A07	Electrical /Electronic Property Problem		
									A0709	Device Sensing Problem		
5309	2017	TBD	35223	Endocardial/interventricular septal pacing lead	CT1684	Cardiac pacemakers and associated devices			A26	Insufficient Device Problem Information		
5310	2018	All other reportable incidents	47270	Cardiac resynchronization therapy implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	F1905	Device Revision or Replacement	A07	Electrical /Electronic Property Problem		
									A0708	Power Problem		
5311	2018	All other reportable incidents			CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A07	Electrical /Electronic Property Problem		
									A0708	Power Problem		
5312	2018	TBD	47270	Cardiac resynchronization therapy implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	F1905	Device Revision or Replacement	A07	Electrical /Electronic Property Problem	01/02/2018	
									A0708	Power Problem	01/02/2018	
5313	2018	All other reportable incidents	47270	Cardiac resynchronization therapy implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A07	Electrical /Electronic Property Problem		
									A0709	Device Sensing Problem		
5314	2017	All other reportable incidents			CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A07	Electrical /Electronic Property Problem		
									A0708	Power Problem		
5315	2017	All other reportable incidents			CT257	Implantable cardiac pacemakers	F1905	Device Revision or Replacement	A07	Electrical /Electronic Property Problem	23/08/2017	
									A0708	Power Problem	23/08/2017	
5316	2017	All other reportable incidents			CT257	Implantable cardiac pacemakers	F1905	Device Revision or Replacement	A07	Electrical /Electronic Property Problem	02/08/2017	
									A0708	Power Problem	02/08/2017	
5317	2018	All other reportable incidents	47270	Cardiac resynchronization therapy implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A07	Electrical /Electronic Property Problem	26/10/2018	
									A0708	Power Problem	26/10/2018	
5318	2018	All other reportable incidents	47270	Cardiac resynchronization therapy implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A07	Electrical /Electronic Property Problem		
									A0708	Power Problem		
5319	2018	All other reportable incidents	35852	Single-chamber implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A07	Electrical /Electronic Property Problem		
									A0709	Device Sensing Problem		
5320	2017	TBD	35223	Endocardial/interventricular septal pacing lead	CT1684	Cardiac pacemakers and associated devices	F1905	Device Revision or Replacement	A0104	Migration or Expulsion of Device		
									A07	Electrical /Electronic Property Problem		
									A0709	Device Sensing Problem		
5321	2017	TBD	35223	Endocardial/interventricular septal pacing lead	CT1684	Cardiac pacemakers and associated devices	E2114	Perforation	A15	Activation, Positioning or Separation Problem		
									A150203	Difficult or Delayed Positioning		
5322	2017	TBD	35223	Endocardial/interventricular septal pacing lead	CT1684	Cardiac pacemakers and associated devices	F1905	Device Revision or Replacement	A0104	Migration or Expulsion of Device		
5323	2018	TBD			CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A09	Output Problem		
									A0908	Incorrect, Inadequate or Imprecise Result or Readings		
5324	2017	TBD			CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A07	Electrical /Electronic Property Problem		
									A0709	Device Sensing Problem		
5325	2018	TBD	35223	Endocardial/interventricular septal pacing lead	CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A15	Activation, Positioning or Separation Problem		
									A150203	Difficult or Delayed Positioning		
5326	2018	All other reportable incidents	35223	Endocardial/interventricular septal pacing lead	CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A07	Electrical /Electronic Property Problem		
									A0709	Device Sensing Problem		
5327	2018	TBD	47270	Cardiac resynchronization therapy implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	F1905	Device Revision or Replacement	A07	Electrical /Electronic Property Problem		
									A0709	Device Sensing Problem		
5328	2018	All other reportable incidents	47270	Cardiac resynchronization therapy implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E1906	Unspecified Infection	A24	No Apparent Device Problem		
									A26	Insufficient Device Problem Information		
5329	2018	Unanticipated serious deterioration in state of health			CT257	Implantable cardiac pacemakers			A26	Insufficient Device Problem Information		
5330	2018	All other reportable incidents	35223	Endocardial/interventricular septal pacing lead	CT1684	Cardiac pacemakers and associated devices	F1905	Device Revision or Replacement	A0104	Migration or Expulsion of Device		
5331	2018	All other reportable incidents	35223	Endocardial/interventricular septal pacing lead	CT1684	Cardiac pacemakers and associated devices	F15	Recognized Device or Procedural Complication	A26	Insufficient Device Problem Information		
5332	2018	Unanticipated serious deterioration in state of health	47270	Cardiac resynchronization therapy implantable defibrillator	CT2866	Implantable cardiac electrical stimulators			A07	Electrical /Electronic Property Problem		
									A0708	Power Problem		
									A11	Computer Software Problem		
									A1102	Application Program Problem		
5333	2017	All other reportable incidents	47270	Cardiac resynchronization therapy implantable	CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A07	Electrical /Electronic Property Problem		

				defibrillator					A0708	Power Problem		
5334	2018	All other reportable incidents	35852	Single-chamber implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A09	Output Problem		
									A090404	Output Above Specifications		
5335	2018	Death			CT1684	Cardiac pacemakers and associated devices	F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A09	Output Problem		
									A0907	No Device Output		
5336	2018	Death	35852	Single-chamber implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	F21	Unexpected Deterioration	A13	Communication or Transmission Problem	17/01/2018	
									A26	Insufficient Device Problem Information	17/01/2018	
5337	2017	All other reportable incidents			CT257	Implantable cardiac pacemakers	F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A07	Electrical /Electronic Property Problem	28/02/2017	
									A0709	Device Sensing Problem	28/02/2017	
5338	2018	Unanticipated serious deterioration in state of health	47270	Cardiac resynchronization therapy implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	F1905	Device Revision or Replacement	A07	Electrical /Electronic Property Problem		
									A0708	Power Problem		
5339	2018	TBD			CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A12	Connection Problem		
5340	2017	TBD	35223	Endocardial/interventricular septal pacing lead	CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A15	Activation, Positioning or Separation Problem		
									A150203	Difficult or Delayed Positioning		
5341	2018	TBD	35852	Single-chamber implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A07	Electrical /Electronic Property Problem		
									A0709	Device Sensing Problem		
5342	2017	All other reportable incidents	35852	Single-chamber implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A1104	Computer Operating System Problem		
5343	2018	TBD			CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A07	Electrical /Electronic Property Problem		
									A0709	Device Sensing Problem		
5344	2018	All other reportable incidents			CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A07	Electrical /Electronic Property Problem		
									A0708	Power Problem		
5345	2017	TBD			CT257	Implantable cardiac pacemakers	F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A07	Electrical /Electronic Property Problem		
									A0708	Power Problem		
5346	2018	All other reportable incidents	47263	Cardiac resynchronization therapy implantable pacemaker	CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A07	Electrical /Electronic Property Problem		
									A0709	Device Sensing Problem		
					CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A07	Electrical /Electronic Property Problem		
									A0709	Device Sensing Problem		
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A07	Electrical /Electronic Property Problem		
									A0709	Device Sensing Problem		
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A07	Electrical /Electronic Property Problem		
									A0709	Device Sensing Problem		
5347	2018	All other reportable incidents			CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A07	Electrical /Electronic Property Problem		
									A0708	Power Problem		
5348	2017	All other reportable incidents			CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A07	Electrical /Electronic Property Problem		
									A0709	Device Sensing Problem		
5349	2017	All other reportable incidents	35223	Endocardial/interventricular septal pacing lead	CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A0104	Migration or Expulsion of Device		
5350	2017	TBD	35223	Endocardial/interventricular septal pacing lead	CT1684	Cardiac pacemakers and associated devices	F1905	Device Revision or Replacement	A05	Mechanical Problem		
									A0510	Retraction Problem		
5351	2017	All other reportable incidents			CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A07	Electrical /Electronic Property Problem		
									A0709	Device Sensing Problem		
5352	2017	All other reportable incidents			CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A07	Electrical /Electronic Property Problem		
									A0708	Power Problem		
5353	2018	TBD			CT257	Implantable cardiac pacemakers			A26	Insufficient Device Problem Information		
5354	2018	All other reportable incidents	47270	Cardiac resynchronization therapy implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A07	Electrical /Electronic Property Problem		
									A0709	Device Sensing Problem		
5355	2018	TBD	47270	Cardiac resynchronization therapy implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	F1905	Device Revision or Replacement	A07	Electrical /Electronic Property Problem		
									A0708	Power Problem		
5356	2018	TBD	35223	Endocardial/interventricular septal pacing lead	CT1684	Cardiac pacemakers and associated devices			A26	Insufficient Device Problem Information		
5357	2017	TBD	35223	Endocardial/interventricular septal pacing lead	CT1684	Cardiac pacemakers and associated devices	F14	Prolonged Episode of Care	A26	Insufficient Device Problem Information		
5358	2017	All other reportable incidents			CT257	Implantable cardiac pacemakers	F1905	Device Revision or Replacement	A07	Electrical /Electronic Property Problem		
									A0708	Power Problem		
5359	2018	All other reportable incidents	35223	Endocardial/interventricular septal pacing lead	CT1684	Cardiac pacemakers and associated devices	F1905	Device Revision or Replacement	A05	Mechanical Problem		
									A051201	Device Dislodged or Dislocated		
5360	2018	All other reportable incidents	35223	Endocardial/interventricular septal pacing lead	CT1684	Cardiac pacemakers and associated devices	F1905	Device Revision or Replacement	A0104	Migration or Expulsion of Device		
5361	2017	TBD	35223	Endocardial/interventricular septal pacing lead	CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A04	Material Integrity Problem	29/09/2016	
									A0405	Degraded	29/09/2016	
5362	2018	All other reportable incidents			CT257	Implantable cardiac pacemakers	F1905	Device Revision or Replacement	A07	Electrical /Electronic Property Problem		
									A0708	Power Problem		
5363	2017	All other reportable incidents			CT257	Implantable cardiac pacemakers	F1905	Device Revision or Replacement	A07	Electrical /Electronic Property Problem		
									A0708	Power Problem		
5364	2017	All other reportable incidents	35223	Endocardial/interventricular septal pacing lead	CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A04	Material Integrity Problem		
									A0404	Crack		
5365	2017	All other reportable incidents			CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A07	Electrical /Electronic Property Problem		
									A0709	Device Sensing Problem		
5366	2018	All other reportable incidents			CT1684	Cardiac pacemakers and associated devices	F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A09	Output Problem		
									A0907	No Device Output		
5367	2017	All other reportable incidents	35223	Endocardial/interventricular septal pacing lead	CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A05	Mechanical Problem		
									A0501	Detachment of Device or Device Component		
5368	2018	All other reportable incidents	47263	Cardiac resynchronization therapy implantable pacemaker	CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A11	Computer Software Problem		
									A1102	Application Program Problem		
					CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A11	Computer Software Problem		
									A1102	Application Program Problem		
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A11	Computer Software Problem		
									A1102	Application Program Problem		
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A11	Computer Software Problem		
									A1102	Application Program Problem		
5369	2017	TBD	35223	Endocardial/interventricular septal pacing lead	CT1684	Cardiac pacemakers and associated devices	F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A0104	Migration or Expulsion of Device		
5370	2017	All other reportable incidents			CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A07	Electrical /Electronic Property Problem	17/08/2017	
									A0708	Power Problem	17/08/2017	
5371	2017	TBD	35852	Single-chamber implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A07	Electrical /Electronic Property Problem		
									A0709	Device Sensing Problem		
5372	2018	All other reportable incidents			CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A09	Output Problem		
									A0907	No Device Output		
5373	2017	TBD	47270	Cardiac resynchronization therapy implantable	CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A07	Electrical /Electronic Property Problem		

				defibrillator					A0709	Device Sensing Problem		
5374	2017	All other reportable incidents			CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A07	Electrical /Electronic Property Problem		
									A0708	Power Problem		
5375	2018	All other reportable incidents	47270	Cardiac resynchronization therapy implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A07	Electrical /Electronic Property Problem		
									A0709	Device Sensing Problem		
5376	2018	TBD	35852	Single-chamber implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A09	Output Problem		
									A0908	Incorrect, Inadequate or Imprecise Result or Readings		
5377	2018	All other reportable incidents			CT257	Implantable cardiac pacemakers	F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A07	Electrical /Electronic Property Problem		
									A0708	Power Problem		
5378	2017	Death	47263	Cardiac resynchronization therapy implantable pacemaker	CT1684	Cardiac pacemakers and associated devices	F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A07	Electrical /Electronic Property Problem	15/08/2016	
									A0709	Device Sensing Problem	15/08/2016	
							F21	Unexpected Deterioration	A07	Electrical /Electronic Property Problem	15/08/2016	
									A0709	Device Sensing Problem	15/08/2016	
					CT1685	Cardiac pacemakers	F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A07	Electrical /Electronic Property Problem	15/08/2016	
									A0709	Device Sensing Problem	15/08/2016	
							F21	Unexpected Deterioration	A07	Electrical /Electronic Property Problem	15/08/2016	
									A0709	Device Sensing Problem	15/08/2016	
					CT257	Implantable cardiac pacemakers	F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A07	Electrical /Electronic Property Problem	15/08/2016	
									A0709	Device Sensing Problem	15/08/2016	
							F21	Unexpected Deterioration	A07	Electrical /Electronic Property Problem	15/08/2016	
									A0709	Device Sensing Problem	15/08/2016	
					CT2866	Implantable cardiac electrical stimulators	F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A07	Electrical /Electronic Property Problem	15/08/2016	
									A0709	Device Sensing Problem	15/08/2016	
							F21	Unexpected Deterioration	A07	Electrical /Electronic Property Problem	15/08/2016	
									A0709	Device Sensing Problem	15/08/2016	
5379	2017	All other reportable incidents	47270	Cardiac resynchronization therapy implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	F1905	Device Revision or Replacement	A07	Electrical /Electronic Property Problem	21/03/2017	
									A0708	Power Problem	21/03/2017	
5380	2018	All other reportable incidents			CT257	Implantable cardiac pacemakers	F1905	Device Revision or Replacement	A07	Electrical /Electronic Property Problem	23/10/2018	
									A0708	Power Problem	23/10/2018	
5381	2018	All other reportable incidents	35223	Endocardial/interventricular septal pacing lead	CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A05	Mechanical Problem	09/10/2018	
									A0510	Retraction Problem	09/10/2018	
5382	2018	All other reportable incidents			CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A07	Electrical /Electronic Property Problem		
									A0708	Power Problem		
5383	2018	All other reportable incidents	47270	Cardiac resynchronization therapy implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A07	Electrical /Electronic Property Problem		
									A0709	Device Sensing Problem		
									A070909	Oversensing		
5384	2017	TBD	35223	Endocardial/interventricular septal pacing lead	CT1684	Cardiac pacemakers and associated devices			A26	Insufficient Device Problem Information		
5385	2017	All other reportable incidents	35223	Endocardial/interventricular septal pacing lead	CT1684	Cardiac pacemakers and associated devices	E2114	Perforation	A26	Insufficient Device Problem Information		
							F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A26	Insufficient Device Problem Information		
5386	2017	All other reportable incidents	35223	Endocardial/interventricular septal pacing lead	CT1684	Cardiac pacemakers and associated devices	F1905	Device Revision or Replacement	A05	Mechanical Problem		
									A051201	Device Dislodged or Dislocated		
5387	2018	All other reportable incidents	35852	Single-chamber implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A07	Electrical /Electronic Property Problem		
									A0709	Device Sensing Problem		
5388	2017	TBD	35223	Endocardial/interventricular septal pacing lead	CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A26	Insufficient Device Problem Information		
5389	2018	All other reportable incidents	35223	Endocardial/interventricular septal pacing lead	CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A05	Mechanical Problem	01/10/2018	
									A0510	Retraction Problem	01/10/2018	
5390	2018	All other reportable incidents	47270	Cardiac resynchronization therapy implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A07	Electrical /Electronic Property Problem		
									A0708	Power Problem		
5391	2018	All other reportable incidents			CT257	Implantable cardiac pacemakers	F1905	Device Revision or Replacement	A07	Electrical /Electronic Property Problem		
									A0708	Power Problem		
5392	2018	All other reportable incidents	47270	Cardiac resynchronization therapy implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	F1905	Device Revision or Replacement	A01	Patient Device Interaction Problem	10/01/2018	
									A17	Compatibility Problem	10/01/2018	
5393	2018	All other reportable incidents	47270	Cardiac resynchronization therapy implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	F1905	Device Revision or Replacement	A07	Electrical /Electronic Property Problem	08/01/2018	
									A0708	Power Problem	08/01/2018	
5394	2018	All other reportable incidents			CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A07	Electrical /Electronic Property Problem	26/01/2018	
									A0708	Power Problem	26/01/2018	
5395	2017	TBD	35223	Endocardial/interventricular septal pacing lead	CT1684	Cardiac pacemakers and associated devices	F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A07	Electrical /Electronic Property Problem		
									A0709	Device Sensing Problem		
5396	2017	TBD	35852	Single-chamber implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E2104	Electric Shock	A07	Electrical /Electronic Property Problem		
									A0709	Device Sensing Problem		
5397	2017	All other reportable incidents	47270	Cardiac resynchronization therapy implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A07	Electrical /Electronic Property Problem		
									A0708	Power Problem		
5398	2018	All other reportable incidents			CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A07	Electrical /Electronic Property Problem		
									A0709	Device Sensing Problem		
5399	2017	TBD			CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A07	Electrical /Electronic Property Problem		
									A0708	Power Problem		
5400	2018	All other reportable incidents			CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A09	Output Problem		
									A0907	No Device Output		
5401	2018	All other reportable incidents	35852	Single-chamber implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	F1905	Device Revision or Replacement	A13	Communication or Transmission Problem	15/06/2018	
									A26	Insufficient Device Problem Information	15/06/2018	
5402	2017	All other reportable incidents	47270	Cardiac resynchronization therapy implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A07	Electrical /Electronic Property Problem		
									A0709	Device Sensing Problem		
5403	2017	All other reportable incidents	47270	Cardiac resynchronization therapy implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A07	Electrical /Electronic Property Problem		
									A0708	Power Problem		
5404	2018	TBD			CT257	Implantable cardiac pacemakers			A26	Insufficient Device Problem Information		
5405	2018	All other reportable incidents			CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A07	Electrical /Electronic Property Problem		
									A0709	Device Sensing Problem		
5406	2018	All other reportable incidents	35223	Endocardial/interventricular septal pacing lead	CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A05	Mechanical Problem		
									A051201	Device Dislodged or Dislocated		
5407	2018	All other reportable incidents	35223	Endocardial/interventricular septal pacing lead	CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A05	Mechanical Problem		
									A0510	Retraction Problem		
5408	2018	All other reportable incidents	35852	Single-chamber implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A07	Electrical /Electronic Property Problem		
									A0709	Device Sensing Problem		
5409	2018	All other reportable incidents			CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A07	Electrical /Electronic Property Problem		
									A0709	Device Sensing Problem		

5410	2017	All other reportable incidents			CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A09	Output Problem		
									A0907	No Device Output		
5411	2018	All other reportable incidents	47270	Cardiac resynchronization therapy implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	F1905	Device Revision or Replacement	A07	Electrical /Electronic Property Problem	20/06/2018	
									A0708	Power Problem	20/06/2018	
5412	2017	All other reportable incidents			CT257	Implantable cardiac pacemakers	F1905	Device Revision or Replacement	A07	Electrical /Electronic Property Problem		
									A0708	Power Problem		
5413	2017	TBD	47270	Cardiac resynchronization therapy implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A11	Computer Software Problem		
									A1102	Application Program Problem		
5414	2017	TBD	35223	Endocardial/interventricular septal pacing lead	CT1684	Cardiac pacemakers and associated devices	F1905	Device Revision or Replacement	A07	Electrical /Electronic Property Problem		
									A0709	Device Sensing Problem		
5415	2017	All other reportable incidents	47270	Cardiac resynchronization therapy implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	F1905	Device Revision or Replacement	A07	Electrical /Electronic Property Problem	03/04/2017	
									A0708	Power Problem	03/04/2017	
5416	2017	All other reportable incidents	35852	Single-chamber implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A09	Output Problem		
									A090503	Radiation Underexposure		
									A0908	Incorrect, Inadequate or Imprecise Result or Readings		
									A1104	Computer Operating System Problem		
5417	2018	Death			CT257	Implantable cardiac pacemakers	F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A15	Activation, Positioning or Separation Problem		
									A150203	Difficult or Delayed Positioning		
5418	2018	All other reportable incidents	47270	Cardiac resynchronization therapy implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	F1905	Device Revision or Replacement	A07	Electrical /Electronic Property Problem	10/01/2018	
									A0708	Power Problem	10/01/2018	
5419	2017	All other reportable incidents	35852	Single-chamber implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A09	Output Problem		
									A0908	Incorrect, Inadequate or Imprecise Result or Readings		
5420	2017	TBD	35223	Endocardial/interventricular septal pacing lead	CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A0104	Migration or Expulsion of Device		
5421	2017	All other reportable incidents	35223	Endocardial/interventricular septal pacing lead	CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A15	Activation, Positioning or Separation Problem		
									A150203	Difficult or Delayed Positioning		
5422	2017	All other reportable incidents			CT257	Implantable cardiac pacemakers	F1905	Device Revision or Replacement	A07	Electrical /Electronic Property Problem	07/03/2017	
									A0709	Device Sensing Problem	07/03/2017	
5423	2018	All other reportable incidents			CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A09	Output Problem		
									A0907	No Device Output		
5424	2018	TBD	35223	Endocardial/interventricular septal pacing lead	CT1684	Cardiac pacemakers and associated devices	F1905	Device Revision or Replacement	A05	Mechanical Problem		
									A0510	Retraction Problem		
									A07	Electrical /Electronic Property Problem		
									A0709	Device Sensing Problem		
									A15	Activation, Positioning or Separation Problem		
									A150301	Separation Failure		
5425	2018	TBD	47263	Cardiac resynchronization therapy implantable pacemaker	CT1684	Cardiac pacemakers and associated devices	F21	Unexpected Deterioration	A07	Electrical /Electronic Property Problem	01/01/2018	
									A0709	Device Sensing Problem	01/01/2018	
					CT1685	Cardiac pacemakers	F21	Unexpected Deterioration	A07	Electrical /Electronic Property Problem	01/01/2018	
									A0709	Device Sensing Problem	01/01/2018	
					CT257	Implantable cardiac pacemakers	F21	Unexpected Deterioration	A07	Electrical /Electronic Property Problem	01/01/2018	
									A0709	Device Sensing Problem	01/01/2018	
					CT2866	Implantable cardiac electrical stimulators	F21	Unexpected Deterioration	A07	Electrical /Electronic Property Problem	01/01/2018	
									A0709	Device Sensing Problem	01/01/2018	
5426	2018	All other reportable incidents	35223	Endocardial/interventricular septal pacing lead	CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A07	Electrical /Electronic Property Problem		
									A0709	Device Sensing Problem		
5427	2018	All other reportable incidents	47270	Cardiac resynchronization therapy implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A07	Electrical /Electronic Property Problem		
									A0708	Power Problem		
5428	2018	All other reportable incidents			CT257	Implantable cardiac pacemakers	F1905	Device Revision or Replacement	A07	Electrical /Electronic Property Problem	26/03/2018	
									A0708	Power Problem	26/03/2018	
									A0902	Display or Visual Feedback Problem	26/03/2018	
									A26	Insufficient Device Problem Information	26/03/2018	
5429	2018	All other reportable incidents			CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A07	Electrical /Electronic Property Problem		
									A0709	Device Sensing Problem		
5430	2017	All other reportable incidents	47270	Cardiac resynchronization therapy implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	F1905	Device Revision or Replacement	A07	Electrical /Electronic Property Problem		
									A0709	Device Sensing Problem		
5431	2017	All other reportable incidents			CT257	Implantable cardiac pacemakers	F15	Recognized Device or Procedural Complication	A01	Patient Device Interaction Problem		
									A17	Compatibility Problem		
5432	2018	Unanticipated serious deterioration in state of health	35223	Endocardial/interventricular septal pacing lead	CT1684	Cardiac pacemakers and associated devices	F21	Unexpected Deterioration	A26	Insufficient Device Problem Information		
5433	2017	All other reportable incidents	47270	Cardiac resynchronization therapy implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A07	Electrical /Electronic Property Problem		
									A0709	Device Sensing Problem		
5434	2018	All other reportable incidents	35852	Single-chamber implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A07	Electrical /Electronic Property Problem		
									A0709	Device Sensing Problem		
									A09	Output Problem		
									A090404	Output Above Specifications		
5435	2018	All other reportable incidents	35852	Single-chamber implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A07	Electrical /Electronic Property Problem		
									A0709	Device Sensing Problem		
5436	2017	All other reportable incidents	35223	Endocardial/interventricular septal pacing lead	CT1684	Cardiac pacemakers and associated devices	F21	Unexpected Deterioration	A26	Insufficient Device Problem Information		
5437	2017	TBD	35852	Single-chamber implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	F1905	Device Revision or Replacement	A07	Electrical /Electronic Property Problem		
									A0708	Power Problem		
5438	2017	TBD	47270	Cardiac resynchronization therapy implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E2401	Insufficient Information	A07	Electrical /Electronic Property Problem		
									A0708	Power Problem		
5439	2017	TBD			CT257	Implantable cardiac pacemakers	F1905	Device Revision or Replacement	A26	Insufficient Device Problem Information		
5440	2017	TBD	35223	Endocardial/interventricular septal pacing lead	CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A07	Electrical /Electronic Property Problem		
									A0709	Device Sensing Problem		
5441	2017	TBD			CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A07	Electrical /Electronic Property Problem		
									A0709	Device Sensing Problem		
5442	2017	All other reportable incidents			CT257	Implantable cardiac pacemakers	F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A11	Computer Software Problem		
									A1102	Application Program Problem		
5443	2017	TBD	35852	Single-chamber implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A07	Electrical /Electronic Property Problem		
									A0709	Device Sensing Problem		
5444	2018	All other reportable incidents	47270	Cardiac resynchronization therapy implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A07	Electrical /Electronic Property Problem		
									A0709	Device Sensing Problem		
5445	2017	TBD	47270	Cardiac resynchronization therapy implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	F1905	Device Revision or Replacement	A07	Electrical /Electronic Property Problem		
									A0708	Power Problem		

5446	2018	Unanticipated serious deterioration in state of health			CT257	Implantable cardiac pacemakers	E2401	Insufficient Information	A26	Insufficient Device Problem Information		
5447	2018	All other reportable incidents	47263	Cardiac resynchronization therapy implantable pacemaker	CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A07	Electrical /Electronic Property Problem		
									A0709	Device Sensing Problem		
					CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A07	Electrical /Electronic Property Problem		
									A0709	Device Sensing Problem		
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A07	Electrical /Electronic Property Problem		
									A0709	Device Sensing Problem		
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A07	Electrical /Electronic Property Problem		
									A0709	Device Sensing Problem		
5448	2018	All other reportable incidents	35223	Endocardial/interventricular septal pacing lead	CT1684	Cardiac pacemakers and associated devices	F1905	Device Revision or Replacement	A07	Electrical /Electronic Property Problem	07/11/2017	
									A0709	Device Sensing Problem	07/11/2017	
5449	2018	All other reportable incidents	47270	Cardiac resynchronization therapy implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	F1905	Device Revision or Replacement	A07	Electrical /Electronic Property Problem		
									A0708	Power Problem		
5450	2018	All other reportable incidents	47270	Cardiac resynchronization therapy implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	F1905	Device Revision or Replacement	A07	Electrical /Electronic Property Problem	22/06/2018	
									A0708	Power Problem	22/06/2018	
5451	2018	All other reportable incidents	47270	Cardiac resynchronization therapy implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	F1905	Device Revision or Replacement	A07	Electrical /Electronic Property Problem	24/10/2018	
									A0708	Power Problem	24/10/2018	
5452	2018	All other reportable incidents	47270	Cardiac resynchronization therapy implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A07	Electrical /Electronic Property Problem	18/01/2018	
									A0708	Power Problem	18/01/2018	
5453	2018	All other reportable incidents	35223	Endocardial/interventricular septal pacing lead	CT1684	Cardiac pacemakers and associated devices	F1905	Device Revision or Replacement	A04	Material Integrity Problem	28/12/2016	
									A0404	Crack	28/12/2016	
5454	2018	All other reportable incidents	47270	Cardiac resynchronization therapy implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A07	Electrical /Electronic Property Problem		
									A0708	Power Problem		
5455	2018	All other reportable incidents	47270	Cardiac resynchronization therapy implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	F1905	Device Revision or Replacement	A07	Electrical /Electronic Property Problem	04/07/2018	
									A0708	Power Problem	04/07/2018	
5456	2018	All other reportable incidents	35852	Single-chamber implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A07	Electrical /Electronic Property Problem		
									A0709	Device Sensing Problem		
5457	2018	All other reportable incidents	47270	Cardiac resynchronization therapy implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E2104	Electric Shock	A09	Output Problem		
									A090404	Output Above Specifications		
5458	2018	All other reportable incidents	35852	Single-chamber implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A07	Electrical /Electronic Property Problem	08/11/2018	08/11/2018
									A0705	Battery Problem	08/11/2018	08/11/2018
									A070504	Premature Discharge of Battery	08/11/2018	08/11/2018
									A0708	Power Problem	08/11/2018	08/11/2018
									A0709	Device Sensing Problem	08/11/2018	08/11/2018
									A072202	Low Impedance	08/11/2018	08/11/2018
							F1905	Device Revision or Replacement	A07	Electrical /Electronic Property Problem	08/11/2018	08/11/2018
									A0705	Battery Problem	08/11/2018	08/11/2018
									A070504	Premature Discharge of Battery	08/11/2018	08/11/2018
									A0708	Power Problem	08/11/2018	08/11/2018
									A0709	Device Sensing Problem	08/11/2018	08/11/2018
									A072202	Low Impedance	08/11/2018	08/11/2018
							F26	No Health Consequences or Impact	A07	Electrical /Electronic Property Problem	08/11/2018	08/11/2018
									A0705	Battery Problem	08/11/2018	08/11/2018
									A070504	Premature Discharge of Battery	08/11/2018	08/11/2018
									A0708	Power Problem	08/11/2018	08/11/2018
									A0709	Device Sensing Problem	08/11/2018	08/11/2018
									A072202	Low Impedance	08/11/2018	08/11/2018
5459	2018	TBD	35852	Single-chamber implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A07	Electrical /Electronic Property Problem		
									A0709	Device Sensing Problem		
5460	2017	All other reportable incidents	35223	Endocardial/interventricular septal pacing lead	CT1684	Cardiac pacemakers and associated devices	F1905	Device Revision or Replacement	A04	Material Integrity Problem	03/08/2017	
									A0404	Crack	03/08/2017	
									A07	Electrical /Electronic Property Problem	03/08/2017	
									A0709	Device Sensing Problem	03/08/2017	
5461	2017	All other reportable incidents			CT257	Implantable cardiac pacemakers	F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A07	Electrical /Electronic Property Problem		
									A0709	Device Sensing Problem		
5462	2018	All other reportable incidents	35852	Single-chamber implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E2007	Fall	A07	Electrical /Electronic Property Problem		
									A0709	Device Sensing Problem		
									A09	Output Problem		
									A090404	Output Above Specifications		
							F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A07	Electrical /Electronic Property Problem		
									A0709	Device Sensing Problem		
									A09	Output Problem		
									A090404	Output Above Specifications		
5463	2017	All other reportable incidents			CT1684	Cardiac pacemakers and associated devices	E2401	Insufficient Information	A09	Output Problem		
									A0907	No Device Output		
5464	2018	All other reportable incidents	35223	Endocardial/interventricular septal pacing lead	CT1684	Cardiac pacemakers and associated devices	F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A09	Output Problem		
									A090404	Output Above Specifications		
5465	2018	Death			CT257	Implantable cardiac pacemakers	F21	Unexpected Deterioration	A26	Insufficient Device Problem Information	06/06/2017	
5466	2018	All other reportable incidents	35852	Single-chamber implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A07	Electrical /Electronic Property Problem		
									A0709	Device Sensing Problem		
5467	2018	All other reportable incidents	35223	Endocardial/interventricular septal pacing lead	CT1684	Cardiac pacemakers and associated devices	F14	Prolonged Episode of Care	A07	Electrical /Electronic Property Problem		
									A0709	Device Sensing Problem		
5468	2018	All other reportable incidents	35223	Endocardial/interventricular septal pacing lead	CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A07	Electrical /Electronic Property Problem		
									A0709	Device Sensing Problem		
5469	2018	All other reportable incidents	35852	Single-chamber implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	F1905	Device Revision or Replacement	A07	Electrical /Electronic Property Problem		
									A0709	Device Sensing Problem		
5470	2018	All other reportable incidents			CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A07	Electrical /Electronic Property Problem	18/10/2018	
									A0708	Power Problem	18/10/2018	
5471	2018	All other reportable incidents	47270	Cardiac resynchronization therapy implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A11	Computer Software Problem		
									A1102	Application Program Problem		
									A16	Protective Measures Problem		
									A1602	Fail-Safe Problem		
5472	2017	All other reportable incidents	35223	Endocardial/interventricular septal pacing lead	CT1684	Cardiac pacemakers and associated devices	E2401	Insufficient Information	A26	Insufficient Device Problem Information		
5473	2018	All other reportable incidents			CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A16	Protective Measures Problem		
									A1602	Fail-Safe Problem		
5474	2018	All other reportable incidents			CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A1104	Computer Operating System Problem		

5475	2017	All other reportable incidents			CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A07	Electrical /Electronic Property Problem		
									A0709	Device Sensing Problem		
5476	2018	All other reportable incidents	35852	Single-chamber implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A07	Electrical /Electronic Property Problem		
									A0709	Device Sensing Problem		
5477	2018	All other reportable incidents	35852	Single-chamber implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A07	Electrical /Electronic Property Problem		
									A0709	Device Sensing Problem		
5478	2018	All other reportable incidents	47270	Cardiac resynchronization therapy implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A07	Electrical /Electronic Property Problem		
									A0708	Power Problem		
5479	2018	Death	47270	Cardiac resynchronization therapy implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	F21	Unexpected Deterioration	A07	Electrical /Electronic Property Problem	18/06/2018	
									A0709	Device Sensing Problem	18/06/2018	
5480	2018	All other reportable incidents			CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A07	Electrical /Electronic Property Problem		
									A0709	Device Sensing Problem		
5481	2018	All other reportable incidents	47270	Cardiac resynchronization therapy implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	F1905	Device Revision or Replacement	A07	Electrical /Electronic Property Problem	21/06/2018	
									A0708	Power Problem	21/06/2018	
5482	2018	All other reportable incidents	35852	Single-chamber implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A07	Electrical /Electronic Property Problem	14/11/2018	
									A0709	Device Sensing Problem	14/11/2018	
							F1001	Absence of Treatment	A07	Electrical /Electronic Property Problem	14/11/2018	
									A0709	Device Sensing Problem	14/11/2018	
							F1905	Device Revision or Replacement	A07	Electrical /Electronic Property Problem	14/11/2018	
									A0709	Device Sensing Problem	14/11/2018	
5483	2017	TBD			CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A1104	Computer Operating System Problem		
5484	2018	All other reportable incidents			CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A09	Output Problem		
									A0907	No Device Output		
5485	2017	TBD	35223	Endocardial/interventricular septal pacing lead	CT1684	Cardiac pacemakers and associated devices	F1905	Device Revision or Replacement	A07	Electrical /Electronic Property Problem		
									A0709	Device Sensing Problem		
5486	2017	TBD	35223	Endocardial/interventricular septal pacing lead	CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A07	Electrical /Electronic Property Problem		
									A0709	Device Sensing Problem		
5487	2018	All other reportable incidents			CT257	Implantable cardiac pacemakers	F1905	Device Revision or Replacement	A1304	Telemetry Discrepancy	14/11/2018	
									A26	Insufficient Device Problem Information	14/11/2018	
							F21	Unexpected Deterioration	A1304	Telemetry Discrepancy	14/11/2018	
									A26	Insufficient Device Problem Information	14/11/2018	
5488	2017	TBD	35223	Endocardial/interventricular septal pacing lead	CT1684	Cardiac pacemakers and associated devices	F1905	Device Revision or Replacement	A0104	Migration or Expulsion of Device		
5489	2017	TBD			CT257	Implantable cardiac pacemakers	E2401	Insufficient Information	A26	Insufficient Device Problem Information		
5490	2017	TBD			CT257	Implantable cardiac pacemakers	F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A07	Electrical /Electronic Property Problem		
									A0709	Device Sensing Problem		
5491	2018	All other reportable incidents	35852	Single-chamber implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A07	Electrical /Electronic Property Problem		
									A0709	Device Sensing Problem		
									A09	Output Problem		
									A090404	Output Above Specifications		
							F1905	Device Revision or Replacement	A07	Electrical /Electronic Property Problem		
									A0709	Device Sensing Problem		
									A09	Output Problem		
									A090404	Output Above Specifications		
5492	2018	All other reportable incidents	35223	Endocardial/interventricular septal pacing lead	CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A05	Mechanical Problem		
									A051201	Device Dislodged or Dislocated		
5493	2017	All other reportable incidents	47263	Cardiac resynchronization therapy implantable pacemaker	CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A07	Electrical /Electronic Property Problem		
									A0709	Device Sensing Problem		
					CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A07	Electrical /Electronic Property Problem		
									A0709	Device Sensing Problem		
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A07	Electrical /Electronic Property Problem		
									A0709	Device Sensing Problem		
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A07	Electrical /Electronic Property Problem		
									A0709	Device Sensing Problem		
5494	2017	TBD	35223	Endocardial/interventricular septal pacing lead	CT1684	Cardiac pacemakers and associated devices			A26	Insufficient Device Problem Information		
5495	2018	All other reportable incidents	35223	Endocardial/interventricular septal pacing lead	CT1684	Cardiac pacemakers and associated devices	F1905	Device Revision or Replacement	A05	Mechanical Problem		
									A0506	Mechanical Jam		
5496	2017	All other reportable incidents	47270	Cardiac resynchronization therapy implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A26	Insufficient Device Problem Information		
5497	2017	All other reportable incidents	47270	Cardiac resynchronization therapy implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A07	Electrical /Electronic Property Problem		
									A0708	Power Problem		
5498	2017	All other reportable incidents			CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A07	Electrical /Electronic Property Problem		
									A0708	Power Problem		
									A09	Output Problem		
									A0907	No Device Output		
5499	2017	All other reportable incidents			CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A09	Output Problem		
									A0907	No Device Output		
5500	2017	All other reportable incidents	47263	Cardiac resynchronization therapy implantable pacemaker	CT1684	Cardiac pacemakers and associated devices	F1905	Device Revision or Replacement	A07	Electrical /Electronic Property Problem	23/03/2017	
									A0708	Power Problem	23/03/2017	
					CT1685	Cardiac pacemakers	F1905	Device Revision or Replacement	A07	Electrical /Electronic Property Problem	23/03/2017	
									A0708	Power Problem	23/03/2017	
					CT257	Implantable cardiac pacemakers	F1905	Device Revision or Replacement	A07	Electrical /Electronic Property Problem	23/03/2017	
									A0708	Power Problem	23/03/2017	
					CT2866	Implantable cardiac electrical stimulators	F1905	Device Revision or Replacement	A07	Electrical /Electronic Property Problem	23/03/2017	
									A0708	Power Problem	23/03/2017	
5501	2018	All other reportable incidents			CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A09	Output Problem		
									A0907	No Device Output		
5502	2018	All other reportable incidents	47270	Cardiac resynchronization therapy implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A09	Output Problem		
									A090404	Output Above Specifications		
5503	2018	All other reportable incidents	47270	Cardiac resynchronization therapy implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A07	Electrical /Electronic Property Problem		
									A0709	Device Sensing Problem		
									A070909	Oversensing		
5504	2018	All other reportable incidents			CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A07	Electrical /Electronic Property Problem		
									A0709	Device Sensing Problem		
5505	2018	All other reportable incidents	47270	Cardiac resynchronization therapy implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A07	Electrical /Electronic Property Problem		
									A0709	Device Sensing Problem		
									A070909	Oversensing		

5506	2018	All other reportable incidents	35223	Endocardial/interventricular septal pacing lead	CT1684	Cardiac pacemakers and associated devices	F21	Unexpected Deterioration	A26	Insufficient Device Problem Information		
5507	2018	All other reportable incidents	47270	Cardiac resynchronization therapy implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A07	Electrical /Electronic Property Problem	31/01/2018	
									A0708	Power Problem	31/01/2018	
5508	2018	All other reportable incidents			CT257	Implantable cardiac pacemakers	F1905	Device Revision or Replacement	A07	Electrical /Electronic Property Problem		
									A0708	Power Problem		
5509	2018	All other reportable incidents	35852	Single-chamber implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A07	Electrical /Electronic Property Problem		
									A0709	Device Sensing Problem		
5510	2018	All other reportable incidents	47263	Cardiac resynchronization therapy implantable pacemaker	CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A07	Electrical /Electronic Property Problem		
									A0709	Device Sensing Problem		
					CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A07	Electrical /Electronic Property Problem		
									A0709	Device Sensing Problem		
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A07	Electrical /Electronic Property Problem		
									A0709	Device Sensing Problem		
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A07	Electrical /Electronic Property Problem		
									A0709	Device Sensing Problem		
5511	2017	All other reportable incidents	35223	Endocardial/interventricular septal pacing lead	CT1684	Cardiac pacemakers and associated devices	F1905	Device Revision or Replacement	A07	Electrical /Electronic Property Problem	11/08/2017	
									A0709	Device Sensing Problem	11/08/2017	
5512	2017	TBD			CT257	Implantable cardiac pacemakers	F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A09	Output Problem		
									A0907	No Device Output		
							F1905	Device Revision or Replacement	A09	Output Problem		
									A0907	No Device Output		
5513	2017	All other reportable incidents			CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A09	Output Problem		
									A0907	No Device Output		
5514	2018	Death	47270	Cardiac resynchronization therapy implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E2401	Insufficient Information	A26	Insufficient Device Problem Information		
5515	2018	All other reportable incidents			CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A07	Electrical /Electronic Property Problem		
									A0709	Device Sensing Problem		
5516	2018	All other reportable incidents			CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A07	Electrical /Electronic Property Problem		
									A0709	Device Sensing Problem		
5517	2018	All other reportable incidents			CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A07	Electrical /Electronic Property Problem		
									A0709	Device Sensing Problem		
5518	2018	All other reportable incidents			CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A07	Electrical /Electronic Property Problem		
									A0709	Device Sensing Problem		
5519	2018	All other reportable incidents			CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A07	Electrical /Electronic Property Problem		
									A0709	Device Sensing Problem		
5520	2018	TBD			CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A07	Electrical /Electronic Property Problem	26/10/2017	
									A0709	Device Sensing Problem	26/10/2017	
5521	2018	All other reportable incidents	47270	Cardiac resynchronization therapy implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A07	Electrical /Electronic Property Problem		
									A0709	Device Sensing Problem		
5522	2018	All other reportable incidents	35223	Endocardial/interventricular septal pacing lead	CT1684	Cardiac pacemakers and associated devices	E1906	Unspecified Infection	A01	Patient Device Interaction Problem		
									A17	Compatibility Problem		
5523	2018	All other reportable incidents	35223	Endocardial/interventricular septal pacing lead	CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A15	Activation, Positioning or Separation Problem	30/01/2018	
									A150101	Activation Failure	30/01/2018	
5524	2018	All other reportable incidents	35223	Endocardial/interventricular septal pacing lead	CT1684	Cardiac pacemakers and associated devices	F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A05	Mechanical Problem		
									A0510	Retraction Problem		
5525	2017	All other reportable incidents	35852	Single-chamber implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A07	Electrical /Electronic Property Problem		
									A0709	Device Sensing Problem		
5526	2017	All other reportable incidents			CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A13	Communication or Transmission Problem		
									A26	Insufficient Device Problem Information		
5527	2017	TBD	35852	Single-chamber implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A07	Electrical /Electronic Property Problem		
									A0709	Device Sensing Problem		
5528	2017	All other reportable incidents	35223	Endocardial/interventricular septal pacing lead	CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A07	Electrical /Electronic Property Problem		
									A0709	Device Sensing Problem		
									A15	Activation, Positioning or Separation Problem		
									A150301	Separation Failure		
5529	2017	TBD	35223	Endocardial/interventricular septal pacing lead	CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A0104	Migration or Expulsion of Device		
5530	2017	TBD	35223	Endocardial/interventricular septal pacing lead	CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A05	Mechanical Problem		
									A0510	Retraction Problem		
5531	2017	TBD			CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A07	Electrical /Electronic Property Problem		
									A0708	Power Problem		
5532	2017	All other reportable incidents	35223	Endocardial/interventricular septal pacing lead	CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A0104	Migration or Expulsion of Device		
									A15	Activation, Positioning or Separation Problem		
									A150203	Difficult or Delayed Positioning		
5533	2017	TBD	35223	Endocardial/interventricular septal pacing lead	CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A07	Electrical /Electronic Property Problem		
									A0709	Device Sensing Problem		
5534	2018	All other reportable incidents			CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A07	Electrical /Electronic Property Problem		
									A0708	Power Problem		
5535	2018	All other reportable incidents	47270	Cardiac resynchronization therapy implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	F1905	Device Revision or Replacement	A07	Electrical /Electronic Property Problem	04/07/2018	
									A0708	Power Problem	04/07/2018	
5536	2018	All other reportable incidents			CT257	Implantable cardiac pacemakers	F1905	Device Revision or Replacement	A07	Electrical /Electronic Property Problem	24/10/2018	
									A0708	Power Problem	24/10/2018	
5537	2017	All other reportable incidents			CT257	Implantable cardiac pacemakers	F1905	Device Revision or Replacement	A07	Electrical /Electronic Property Problem	04/04/2017	
									A0708	Power Problem	04/04/2017	
5538	2018	All other reportable incidents	35852	Single-chamber implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	F1905	Device Revision or Replacement	A07	Electrical /Electronic Property Problem	25/10/2018	
									A0708	Power Problem	25/10/2018	
5539	2018	All other reportable incidents	35223	Endocardial/interventricular septal pacing lead	CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A05	Mechanical Problem		
									A0510	Retraction Problem		
5540	2018	TBD			CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A07	Electrical /Electronic Property Problem		
									A0709	Device Sensing Problem		
5541	2018	TBD	35223	Endocardial/interventricular septal pacing lead	CT1684	Cardiac pacemakers and associated devices	F1905	Device Revision or Replacement	A07	Electrical /Electronic Property Problem		
									A0709	Device Sensing Problem		
5542	2017	TBD			CT257	Implantable cardiac pacemakers	F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A11	Computer Software Problem		
									A1102	Application Program Problem		
5543	2018	All other reportable incidents	47270	Cardiac resynchronization therapy implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A07	Electrical /Electronic Property Problem		
									A0708	Power Problem		
5544	2017	TBD			CT257	Implantable cardiac pacemakers	F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A07	Electrical /Electronic Property Problem		

									A0709	Device Sensing Problem		
5545	2017	All other reportable incidents	47270	Cardiac resynchronization therapy implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A07	Electrical /Electronic Property Problem		
									A0708	Power Problem		
									A26	Insufficient Device Problem Information		
5546	2018	All other reportable incidents			CT257	Implantable cardiac pacemakers	F1905	Device Revision or Replacement	A07	Electrical /Electronic Property Problem		
									A0708	Power Problem		
5547	2017	All other reportable incidents	60789	Intracardiac pacemaker	CT130	Catheters and associated devices	F1905	Device Revision or Replacement	A26	Insufficient Device Problem Information		
					CT145	Cardiovascular devices	F1905	Device Revision or Replacement	A26	Insufficient Device Problem Information		
					CT1684	Cardiac pacemakers and associated devices	F1905	Device Revision or Replacement	A26	Insufficient Device Problem Information		
					CT1685	Cardiac pacemakers	F1905	Device Revision or Replacement	A26	Insufficient Device Problem Information		
					CT233	Surgical	F1905	Device Revision or Replacement	A26	Insufficient Device Problem Information		
					CT2406	Implantable	F1905	Device Revision or Replacement	A26	Insufficient Device Problem Information		
					CT2407	Active implantable	F1905	Device Revision or Replacement	A26	Insufficient Device Problem Information		
					CT257	Implantable cardiac pacemakers	F1905	Device Revision or Replacement	A26	Insufficient Device Problem Information		
					CT2727	Active implant devices/ancillaries	F1905	Device Revision or Replacement	A26	Insufficient Device Problem Information		
					CT2750	Implantable cardiac electrical stimulators and associated devices	F1905	Device Revision or Replacement	A26	Insufficient Device Problem Information		
					CT2818	Catheterization/Cannulation/Fluid path devices	F1905	Device Revision or Replacement	A26	Insufficient Device Problem Information		
					CT2866	Implantable cardiac electrical stimulators	F1905	Device Revision or Replacement	A26	Insufficient Device Problem Information		
					CT2985	Active	F1905	Device Revision or Replacement	A26	Insufficient Device Problem Information		
					CT3001	Implantable metal	F1905	Device Revision or Replacement	A26	Insufficient Device Problem Information		
					CT3002	Implantable synthetic polymer	F1905	Device Revision or Replacement	A26	Insufficient Device Problem Information		
					CT3006	Implantable inorganic material	F1905	Device Revision or Replacement	A26	Insufficient Device Problem Information		
					CT314	Electrically powered	F1905	Device Revision or Replacement	A26	Insufficient Device Problem Information		
					CT321	Long-term surgical invasive	F1905	Device Revision or Replacement	A26	Insufficient Device Problem Information		
					CT981	Single use	F1905	Device Revision or Replacement	A26	Insufficient Device Problem Information		
					CT983	Surgical invasive	F1905	Device Revision or Replacement	A26	Insufficient Device Problem Information		
5548	2018	All other reportable incidents	47270	Cardiac resynchronization therapy implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	F1905	Device Revision or Replacement	A07	Electrical /Electronic Property Problem	27/06/2018	
									A0708	Power Problem	27/06/2018	
5549	2017	All other reportable incidents	35223	Endocardial/interventricular septal pacing lead	CT1684	Cardiac pacemakers and associated devices	F1905	Device Revision or Replacement	A04	Material Integrity Problem	31/03/2017	
									A0404	Crack	31/03/2017	
5550	2018	All other reportable incidents	35852	Single-chamber implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A09	Output Problem	22/06/2018	
									A090404	Output Above Specifications	22/06/2018	
5551	2018	All other reportable incidents	35223	Endocardial/interventricular septal pacing lead	CT1684	Cardiac pacemakers and associated devices	F1905	Device Revision or Replacement	A07	Electrical /Electronic Property Problem		
									A0709	Device Sensing Problem		
5552	2017	All other reportable incidents			CT1684	Cardiac pacemakers and associated devices	F05	Delay to Treatment/ Therapy	A1104	Computer Operating System Problem		
5553	2018	All other reportable incidents	35223	Endocardial/interventricular septal pacing lead	CT1684	Cardiac pacemakers and associated devices	F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A07	Electrical /Electronic Property Problem	27/06/2018	
									A0709	Device Sensing Problem	27/06/2018	
5554	2018	All other reportable incidents	35852	Single-chamber implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A07	Electrical /Electronic Property Problem		
									A0709	Device Sensing Problem		
5555	2017	TBD	35223	Endocardial/interventricular septal pacing lead	CT1684	Cardiac pacemakers and associated devices	E2401	Insufficient Information	A26	Insufficient Device Problem Information		
5556	2017	TBD	47270	Cardiac resynchronization therapy implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	F1905	Device Revision or Replacement	A07	Electrical /Electronic Property Problem		
									A0708	Power Problem		
5557	2017	TBD	35852	Single-chamber implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A07	Electrical /Electronic Property Problem		
									A0709	Device Sensing Problem		
5558	2017	TBD	35852	Single-chamber implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A09	Output Problem		
									A090404	Output Above Specifications		
5559	2018	All other reportable incidents	47270	Cardiac resynchronization therapy implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A07	Electrical /Electronic Property Problem	25/10/2018	
									A0708	Power Problem	25/10/2018	
5560	2017	All other reportable incidents			CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A07	Electrical /Electronic Property Problem		
									A0708	Power Problem		
									A11	Computer Software Problem		
									A1102	Application Program Problem		
5561	2017	All other reportable incidents			CT257	Implantable cardiac pacemakers	F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A07	Electrical /Electronic Property Problem	30/03/2017	
									A0708	Power Problem	30/03/2017	
5562	2017	TBD			CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A07	Electrical /Electronic Property Problem		
									A0709	Device Sensing Problem		
5563	2018	TBD	35223	Endocardial/interventricular septal pacing lead	CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A15	Activation, Positioning or Separation Problem		
									A150203	Difficult or Delayed Positioning		
5564	2018	All other reportable incidents	47270	Cardiac resynchronization therapy implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A07	Electrical /Electronic Property Problem	27/06/2018	
									A0708	Power Problem	27/06/2018	
5565	2018	All other reportable incidents	35223	Endocardial/interventricular septal pacing lead	CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A15	Activation, Positioning or Separation Problem	08/11/2017	
									A150203	Difficult or Delayed Positioning	08/11/2017	
5566	2018	TBD			CT257	Implantable cardiac pacemakers	F1905	Device Revision or Replacement	A07	Electrical /Electronic Property Problem	01/01/2018	
									A0709	Device Sensing Problem	01/01/2018	
5567	2018	All other reportable incidents			CT257	Implantable cardiac pacemakers	F1905	Device Revision or Replacement	A07	Electrical /Electronic Property Problem	25/01/2018	
									A0708	Power Problem	25/01/2018	
									A16	Protective Measures Problem	25/01/2018	
									A1602	Fail-Safe Problem	25/01/2018	
5568	2017	TBD			CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A07	Electrical /Electronic Property Problem		
									A0708	Power Problem		
5569	2018	All other reportable incidents			CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A07	Electrical /Electronic Property Problem		
									A0709	Device Sensing Problem		
5570	2018	All other reportable incidents	47270	Cardiac resynchronization therapy implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A07	Electrical /Electronic Property Problem		
									A0709	Device Sensing Problem		
5571	2018	All other reportable incidents	35852	Single-chamber implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A07	Electrical /Electronic Property Problem		
									A0709	Device Sensing Problem		
5572	2017	All other reportable incidents			CT257	Implantable cardiac pacemakers	F1905	Device Revision or Replacement	A13	Communication or Transmission Problem	01/09/2017	
									A26	Insufficient Device Problem Information	01/09/2017	
5573	2017	All other reportable incidents	35223	Endocardial/interventricular septal pacing lead	CT1684	Cardiac pacemakers and associated devices	F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A0104	Migration or Expulsion of Device		
5574	2017	Unanticipated serious deterioration in state of health	47270	Cardiac resynchronization therapy implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E1906	Unspecified Infection	A01	Patient Device Interaction Problem		
									A17	Compatibility Problem		
5575	2018	Unanticipated serious deterioration in state of health			CT257	Implantable cardiac pacemakers	E2401	Insufficient Information	A26	Insufficient Device Problem Information		
5576	2018	All other reportable incidents	47270	Cardiac resynchronization therapy implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	F1905	Device Revision or Replacement	A07	Electrical /Electronic Property Problem		
									A0708	Power Problem		
5577	2018	All other reportable incidents	35223	Endocardial/interventricular septal pacing lead	CT1684	Cardiac pacemakers and associated devices	E2008	Foreign Body In Patient	A04	Material Integrity Problem	24/10/2018	

									A0413	Material Separation	24/10/2018	
							F14	Prolonged Episode of Care	A04	Material Integrity Problem	24/10/2018	
									A0413	Material Separation	24/10/2018	
5578	2018	All other reportable incidents	35223	Endocardial/interventricular septal pacing lead	CT1684	Cardiac pacemakers and associated devices	F1905	Device Revision or Replacement	A07	Electrical /Electronic Property Problem		
									A0709	Device Sensing Problem		
5579	2018	Unanticipated serious deterioration in state of health			CT257	Implantable cardiac pacemakers	F1905	Device Revision or Replacement	A07	Electrical /Electronic Property Problem	22/06/2018	
									A0708	Power Problem	22/06/2018	
5580	2018	All other reportable incidents			CT257	Implantable cardiac pacemakers	F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A1104	Computer Operating System Problem		
5581	2018	All other reportable incidents	35852	Single-chamber implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A07	Electrical /Electronic Property Problem		
									A0709	Device Sensing Problem		
5582	2018	All other reportable incidents			CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A07	Electrical /Electronic Property Problem		
									A0708	Power Problem		
5583	2017	TBD	35223	Endocardial/interventricular septal pacing lead	CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A07	Electrical /Electronic Property Problem		
									A0709	Device Sensing Problem		
5584	2017	TBD	35223	Endocardial/interventricular septal pacing lead	CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A15	Activation, Positioning or Separation Problem		
									A150203	Difficult or Delayed Positioning		
5585	2018	All other reportable incidents			CT257	Implantable cardiac pacemakers	F1905	Device Revision or Replacement	A07	Electrical /Electronic Property Problem	23/10/2018	
									A0708	Power Problem	23/10/2018	
									A0709	Device Sensing Problem	23/10/2018	
5586	2018	TBD	47270	Cardiac resynchronization therapy implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A09	Output Problem		
									A0908	Incorrect, Inadequate or Imprecise Result or Readings		
5587	2018	TBD	47270	Cardiac resynchronization therapy implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	F1905	Device Revision or Replacement	A07	Electrical /Electronic Property Problem		
									A0708	Power Problem		
5588	2018	TBD			CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A07	Electrical /Electronic Property Problem		
									A0709	Device Sensing Problem		
5589	2018	TBD	35223	Endocardial/interventricular septal pacing lead	CT1684	Cardiac pacemakers and associated devices	E0402	Hypersensitivity/Allergic Reaction	A26	Insufficient Device Problem Information		
5590	2017	TBD			CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A09	Output Problem		
									A0907	No Device Output		
5591	2018	TBD	35223	Endocardial/interventricular septal pacing lead	CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A05	Mechanical Problem		
									A0506	Mechanical Jam		
5592	2017	All other reportable incidents			CT257	Implantable cardiac pacemakers	F1905	Device Revision or Replacement	A07	Electrical /Electronic Property Problem	13/04/2017	
									A0708	Power Problem	13/04/2017	
5593	2018	All other reportable incidents	47263	Cardiac resynchronization therapy implantable pacemaker	CT1684	Cardiac pacemakers and associated devices	F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A09	Output Problem		
									A090404	Output Above Specifications		
					CT1685	Cardiac pacemakers	F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A09	Output Problem		
									A090404	Output Above Specifications		
					CT257	Implantable cardiac pacemakers	F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A09	Output Problem		
									A090404	Output Above Specifications		
					CT2866	Implantable cardiac electrical stimulators	F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A09	Output Problem		
									A090404	Output Above Specifications		
5594	2018	Unanticipated serious deterioration in state of health	47270	Cardiac resynchronization therapy implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	F1905	Device Revision or Replacement	A07	Electrical /Electronic Property Problem	22/11/2018	
									A0708	Power Problem	22/11/2018	
5595	2018	All other reportable incidents			CT257	Implantable cardiac pacemakers	F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A07	Electrical /Electronic Property Problem		
									A0712	Pacing Problem		
									A071207	Pacing Inadequately		
5596	2018	All other reportable incidents	47270	Cardiac resynchronization therapy implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A07	Electrical /Electronic Property Problem		
									A0708	Power Problem		
5597	2018	All other reportable incidents			CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A07	Electrical /Electronic Property Problem		
									A0709	Device Sensing Problem		
5598	2018	All other reportable incidents			CT257	Implantable cardiac pacemakers	E0752	Unspecified Respiratory Problem	A11	Computer Software Problem		
									A1102	Application Program Problem		
5599	2017	TBD	47270	Cardiac resynchronization therapy implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A07	Electrical /Electronic Property Problem		
									A0709	Device Sensing Problem		
5600	2018	All other reportable incidents			CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A07	Electrical /Electronic Property Problem		
									A0709	Device Sensing Problem		
5601	2017	All other reportable incidents	35852	Single-chamber implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A07	Electrical /Electronic Property Problem		
									A0709	Device Sensing Problem		
5602	2018	All other reportable incidents	35852	Single-chamber implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	F1905	Device Revision or Replacement	A07	Electrical /Electronic Property Problem	27/11/2018	
									A0708	Power Problem	27/11/2018	
5603	2017	TBD	47270	Cardiac resynchronization therapy implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	F1905	Device Revision or Replacement	A07	Electrical /Electronic Property Problem		
									A0708	Power Problem		
5604	2017	TBD	35223	Endocardial/interventricular septal pacing lead	CT1684	Cardiac pacemakers and associated devices	E2401	Insufficient Information	A26	Insufficient Device Problem Information		
5605	2018	All other reportable incidents			CT1684	Cardiac pacemakers and associated devices	F05	Delay to Treatment/ Therapy	A1104	Computer Operating System Problem		
5606	2017	TBD	35223	Endocardial/interventricular septal pacing lead	CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A05	Mechanical Problem		
									A0506	Mechanical Jam		
5607	2018	All other reportable incidents	35852	Single-chamber implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A07	Electrical /Electronic Property Problem		
									A0709	Device Sensing Problem		
							F1905	Device Revision or Replacement	A07	Electrical /Electronic Property Problem		
									A0709	Device Sensing Problem		
5608	2018	TBD	35223	Endocardial/interventricular septal pacing lead	CT1684	Cardiac pacemakers and associated devices	F1905	Device Revision or Replacement	A04	Material Integrity Problem		
									A0410	Material Perforation		
5609	2018	All other reportable incidents	35852	Single-chamber implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A0902	Display or Visual Feedback Problem		
									A26	Insufficient Device Problem Information		
5610	2017	All other reportable incidents	35852	Single-chamber implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A07	Electrical /Electronic Property Problem		
									A0708	Power Problem		
5611	2018	All other reportable incidents	47270	Cardiac resynchronization therapy implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E0514	Thrombosis/Thrombus	A05	Mechanical Problem	21/11/2017	
									A0501	Detachment of Device or Device Component	21/11/2017	
							F1905	Device Revision or Replacement	A05	Mechanical Problem	21/11/2017	
									A0501	Detachment of Device or Device Component	21/11/2017	
5612	2018	All other reportable incidents	47270	Cardiac resynchronization therapy implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A07	Electrical /Electronic Property Problem		
									A0709	Device Sensing Problem		
5613	2018	All other reportable incidents	35223	Endocardial/interventricular septal pacing lead	CT1684	Cardiac pacemakers and associated devices	F1905	Device Revision or Replacement	A07	Electrical /Electronic Property Problem		
									A0709	Device Sensing Problem		
5614	2018	Unanticipated serious deterioration in state of health			CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A07	Electrical /Electronic Property Problem	21/11/2018	
									A0708	Power Problem	21/11/2018	

5615	2018	All other reportable incidents	35223	Endocardial/interventricular septal pacing lead	CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A05	Mechanical Problem		
									A0510	Retraction Problem		
									A07	Electrical /Electronic Property Problem		
									A0709	Device Sensing Problem		
5616	2018	All other reportable incidents	35223	Endocardial/interventricular septal pacing lead	CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A26	Insufficient Device Problem Information		
5617	2017	TBD			CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A07	Electrical /Electronic Property Problem		
									A0709	Device Sensing Problem		
5618	2018	TBD	35223	Endocardial/interventricular septal pacing lead	CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A15	Activation, Positioning or Separation Problem		
									A150101	Activation Failure		
5619	2018	All other reportable incidents	47270	Cardiac resynchronization therapy implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A07	Electrical /Electronic Property Problem		
									A0708	Power Problem		
5620	2018	All other reportable incidents	35223	Endocardial/interventricular septal pacing lead	CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A05	Mechanical Problem	26/10/2018	
									A0510	Retraction Problem	26/10/2018	
									A07	Electrical /Electronic Property Problem	26/10/2018	
									A0709	Device Sensing Problem	26/10/2018	
5621	2018	All other reportable incidents	35852	Single-chamber implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A09	Output Problem		
									A090404	Output Above Specifications		
5622	2017	TBD	35223	Endocardial/interventricular septal pacing lead	CT1684	Cardiac pacemakers and associated devices	E2401	Insufficient Information	A26	Insufficient Device Problem Information		
5623	2018	All other reportable incidents	35223	Endocardial/interventricular septal pacing lead	CT1684	Cardiac pacemakers and associated devices	F14	Prolonged Episode of Care	A07	Electrical /Electronic Property Problem	23/05/2018	
									A0709	Device Sensing Problem	23/05/2018	
5624	2018	All other reportable incidents			CT1684	Cardiac pacemakers and associated devices	F05	Delay to Treatment/ Therapy	A09	Output Problem		
									A0907	No Device Output		
5625	2017	TBD	35223	Endocardial/interventricular septal pacing lead	CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A0104	Migration or Expulsion of Device		
5626	2017	TBD			CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A1304	Telemetry Discrepancy		
									A26	Insufficient Device Problem Information		
5627	2017	All other reportable incidents			CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A07	Electrical /Electronic Property Problem		
									A0708	Power Problem		
5628	2017	TBD	47270	Cardiac resynchronization therapy implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	F1905	Device Revision or Replacement	A07	Electrical /Electronic Property Problem		
									A0708	Power Problem		
5629	2018	All other reportable incidents			CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A07	Electrical /Electronic Property Problem		
									A0708	Power Problem		
5630	2018	All other reportable incidents			CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A07	Electrical /Electronic Property Problem		
									A0708	Power Problem		
5631	2018	All other reportable incidents	47270	Cardiac resynchronization therapy implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	F1905	Device Revision or Replacement	A26	Insufficient Device Problem Information	25/02/2015	
5632	2018	All other reportable incidents	47270	Cardiac resynchronization therapy implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A11	Computer Software Problem		
									A1102	Application Program Problem		
5633	2018	All other reportable incidents	35852	Single-chamber implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A07	Electrical /Electronic Property Problem		
									A0709	Device Sensing Problem		
5634	2018	All other reportable incidents	35223	Endocardial/interventricular septal pacing lead	CT1684	Cardiac pacemakers and associated devices	F1905	Device Revision or Replacement	A07	Electrical /Electronic Property Problem		
									A0709	Device Sensing Problem		
5635	2017	All other reportable incidents	47270	Cardiac resynchronization therapy implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A07	Electrical /Electronic Property Problem		
									A0709	Device Sensing Problem		
5636	2018	All other reportable incidents	47270	Cardiac resynchronization therapy implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A07	Electrical /Electronic Property Problem		
									A0708	Power Problem		
5637	2018	All other reportable incidents	47270	Cardiac resynchronization therapy implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	F1905	Device Revision or Replacement	A07	Electrical /Electronic Property Problem	05/07/2018	
									A0708	Power Problem	05/07/2018	
5638	2018	All other reportable incidents	47270	Cardiac resynchronization therapy implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A07	Electrical /Electronic Property Problem		
									A0708	Power Problem		
5639	2018	All other reportable incidents	47270	Cardiac resynchronization therapy implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	F1905	Device Revision or Replacement	A07	Electrical /Electronic Property Problem	21/11/2018	
									A0708	Power Problem	21/11/2018	
5640	2017	All other reportable incidents	47270	Cardiac resynchronization therapy implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	F1905	Device Revision or Replacement	A07	Electrical /Electronic Property Problem	12/09/2017	
									A0708	Power Problem	12/09/2017	
5641	2018	All other reportable incidents	35852	Single-chamber implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E1906	Unspecified Infection	A01	Patient Device Interaction Problem	01/11/2018	
									A17	Compatibility Problem	01/11/2018	
							F1905	Device Revision or Replacement	A01	Patient Device Interaction Problem	01/11/2018	
									A17	Compatibility Problem	01/11/2018	
5642	2018	All other reportable incidents	47270	Cardiac resynchronization therapy implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A1104	Computer Operating System Problem	26/06/2018	
									A13	Communication or Transmission Problem	26/06/2018	
									A26	Insufficient Device Problem Information	26/06/2018	
5643	2017	TBD	35223	Endocardial/interventricular septal pacing lead	CT1684	Cardiac pacemakers and associated devices	E2401	Insufficient Information	A26	Insufficient Device Problem Information		
5644	2018	All other reportable incidents	35852	Single-chamber implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A09	Output Problem		
									A090404	Output Above Specifications		
5645	2017	All other reportable incidents	35852	Single-chamber implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A07	Electrical /Electronic Property Problem	02/07/2018	
									A0709	Device Sensing Problem	02/07/2018	
									A09	Output Problem	02/07/2018	
									A090404	Output Above Specifications	02/07/2018	
5646	2018	TBD	35852	Single-chamber implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A09	Output Problem		
									A090404	Output Above Specifications		
5647	2017	TBD			CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A07	Electrical /Electronic Property Problem		
									A0709	Device Sensing Problem		
5648	2017	TBD	47270	Cardiac resynchronization therapy implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A07	Electrical /Electronic Property Problem		
									A0709	Device Sensing Problem		
5649	2018	All other reportable incidents	47270	Cardiac resynchronization therapy implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	F1905	Device Revision or Replacement	A07	Electrical /Electronic Property Problem		
									A0708	Power Problem		
5650	2017	TBD	47270	Cardiac resynchronization therapy implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	F1905	Device Revision or Replacement	A07	Electrical /Electronic Property Problem		
									A0708	Power Problem		
5651	2017	TBD	47270	Cardiac resynchronization therapy implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A07	Electrical /Electronic Property Problem		
									A0708	Power Problem		
5652	2018	TBD			CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A07	Electrical /Electronic Property Problem		
									A0709	Device Sensing Problem		
5653	2018	All other reportable incidents	35223	Endocardial/interventricular septal pacing lead	CT1684	Cardiac pacemakers and associated devices	F14	Prolonged Episode of Care	A0104	Migration or Expulsion of Device		
							F1905	Device Revision or Replacement	A0104	Migration or Expulsion of Device		
5654	2018	All other reportable incidents			CT257	Implantable cardiac pacemakers	F1905	Device Revision or Replacement	A07	Electrical /Electronic Property Problem		
									A0708	Power Problem		

5655	2018	All other reportable incidents	35223	Endocardial/interventricular septal pacing lead	CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A15	Activation, Positioning or Separation Problem	01/11/2018	
									A150203	Difficult or Delayed Positioning	01/11/2018	
5656	2018	All other reportable incidents	47270	Cardiac resynchronization therapy implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	F1905	Device Revision or Replacement	A13	Communication or Transmission Problem	26/06/2018	
									A26	Insufficient Device Problem Information	26/06/2018	
5657	2017	TBD	47270	Cardiac resynchronization therapy implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A07	Electrical /Electronic Property Problem		
									A0708	Power Problem		
5658	2017	All other reportable incidents	47270	Cardiac resynchronization therapy implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	F1905	Device Revision or Replacement	A07	Electrical /Electronic Property Problem	03/05/2017	
									A0708	Power Problem	03/05/2017	
5659	2017	All other reportable incidents			CT257	Implantable cardiac pacemakers	F1905	Device Revision or Replacement	A07	Electrical /Electronic Property Problem	11/09/2017	
									A0708	Power Problem	11/09/2017	
5660	2018	TBD	35223	Endocardial/interventricular septal pacing lead	CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A04	Material Integrity Problem		
									A0405	Degraded		
5661	2018	TBD	35223	Endocardial/interventricular septal pacing lead	CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A15	Activation, Positioning or Separation Problem		
									A150101	Activation Failure		
5662	2018	All other reportable incidents	35223	Endocardial/interventricular septal pacing lead	CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A0104	Migration or Expulsion of Device		
5663	2018	All other reportable incidents	47270	Cardiac resynchronization therapy implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	F1905	Device Revision or Replacement	A07	Electrical /Electronic Property Problem		
									A0708	Power Problem		
5664	2018	All other reportable incidents			CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A07	Electrical /Electronic Property Problem		
									A0709	Device Sensing Problem		
5665	2017	All other reportable incidents			CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A07	Electrical /Electronic Property Problem		
									A0708	Power Problem		
5666	2017	All other reportable incidents			CT257	Implantable cardiac pacemakers	F1905	Device Revision or Replacement	A07	Electrical /Electronic Property Problem	08/09/2017	
									A0708	Power Problem	08/09/2017	
5667	2018	All other reportable incidents	47270	Cardiac resynchronization therapy implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A07	Electrical /Electronic Property Problem		
									A0709	Device Sensing Problem		
5668	2018	All other reportable incidents	47263	Cardiac resynchronization therapy implantable pacemaker	CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A07	Electrical /Electronic Property Problem		
									A0709	Device Sensing Problem		
									A070909	Oversensing		
									A0722	Impedance Problem		
									A072201	High Impedance		
					CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A07	Electrical /Electronic Property Problem		
									A0709	Device Sensing Problem		
									A070909	Oversensing		
									A0722	Impedance Problem		
									A072201	High Impedance		
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A07	Electrical /Electronic Property Problem		
									A0709	Device Sensing Problem		
									A070909	Oversensing		
									A0722	Impedance Problem		
									A072201	High Impedance		
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A07	Electrical /Electronic Property Problem		
									A0709	Device Sensing Problem		
									A070909	Oversensing		
									A0722	Impedance Problem		
									A072201	High Impedance		
5669	2018	All other reportable incidents	35223	Endocardial/interventricular septal pacing lead	CT1684	Cardiac pacemakers and associated devices	F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A07	Electrical /Electronic Property Problem		
									A0709	Device Sensing Problem		
5670	2017	All other reportable incidents	47270	Cardiac resynchronization therapy implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	F1905	Device Revision or Replacement	A07	Electrical /Electronic Property Problem		
									A0708	Power Problem		
5671	2018	All other reportable incidents	35852	Single-chamber implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A13	Communication or Transmission Problem		
									A26	Insufficient Device Problem Information		
5672	2018	TBD	35223	Endocardial/interventricular septal pacing lead	CT1684	Cardiac pacemakers and associated devices			A26	Insufficient Device Problem Information		
5673	2018	Death	35852	Single-chamber implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A07	Electrical /Electronic Property Problem		
									A0709	Device Sensing Problem		
5674	2017	All other reportable incidents	47270	Cardiac resynchronization therapy implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A07	Electrical /Electronic Property Problem		
									A0708	Power Problem		
5675	2018	TBD	35223	Endocardial/interventricular septal pacing lead	CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A05	Mechanical Problem		
									A0506	Mechanical Jam		
5676	2018	Death			CT257	Implantable cardiac pacemakers	F21	Unexpected Deterioration	A26	Insufficient Device Problem Information		
5677	2018	All other reportable incidents	35852	Single-chamber implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A07	Electrical /Electronic Property Problem		
									A0709	Device Sensing Problem		
5678	2017	All other reportable incidents	47270	Cardiac resynchronization therapy implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	F1905	Device Revision or Replacement	A07	Electrical /Electronic Property Problem	21/04/2017	
									A0708	Power Problem	21/04/2017	
5679	2017	All other reportable incidents	35223	Endocardial/interventricular septal pacing lead	CT1684	Cardiac pacemakers and associated devices	E2401	Insufficient Information	A26	Insufficient Device Problem Information		
5680	2018	All other reportable incidents	35852	Single-chamber implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A07	Electrical /Electronic Property Problem		
									A0709	Device Sensing Problem		
5681	2018	All other reportable incidents	47270	Cardiac resynchronization therapy implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A07	Electrical /Electronic Property Problem		
									A0708	Power Problem		
5682	2018	All other reportable incidents			CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A17	Compatibility Problem	10/05/2018	
									A1702	Device-Device Incompatibility	10/05/2018	
5683	2018	All other reportable incidents	35852	Single-chamber implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A07	Electrical /Electronic Property Problem		
									A0709	Device Sensing Problem		
5684	2017	TBD	47270	Cardiac resynchronization therapy implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	F1905	Device Revision or Replacement	A07	Electrical /Electronic Property Problem		
									A0708	Power Problem		
5685	2018	All other reportable incidents	35223	Endocardial/interventricular septal pacing lead	CT1684	Cardiac pacemakers and associated devices	F1905	Device Revision or Replacement	A0104	Migration or Expulsion of Device		
5686	2018	All other reportable incidents	47270	Cardiac resynchronization therapy implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	F05	Delay to Treatment/ Therapy	A07	Electrical /Electronic Property Problem		
									A0708	Power Problem		
5687	2017	TBD	35223	Endocardial/interventricular septal pacing lead	CT1684	Cardiac pacemakers and associated devices	F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A26	Insufficient Device Problem Information		
5688	2018	All other reportable incidents	35223	Endocardial/interventricular septal pacing lead	CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A05	Mechanical Problem		
									A0510	Retraction Problem		
5689	2017	All other reportable incidents	47270	Cardiac resynchronization therapy implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A11	Computer Software Problem		
									A1102	Application Program Problem		
5690	2017	All other reportable incidents	35223	Endocardial/interventricular septal pacing lead	CT1684	Cardiac pacemakers and associated devices	F1905	Device Revision or Replacement	A07	Electrical /Electronic Property Problem	05/09/2017	
									A0709	Device Sensing Problem	05/09/2017	

5691	2018	TBD	35852	Single-chamber implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A0104	Migration or Expulsion of Device		
5692	2018	All other reportable incidents	35852	Single-chamber implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A07	Electrical /Electronic Property Problem		
									A0709	Device Sensing Problem		
5693	2018	All other reportable incidents	35852	Single-chamber implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A07	Electrical /Electronic Property Problem		
									A0709	Device Sensing Problem		
5694	2018	All other reportable incidents	35852	Single-chamber implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	F1905	Device Revision or Replacement	A07	Electrical /Electronic Property Problem		
									A0713	Defibrillation/Stimulation Problem		
									A071301	Failure to Deliver Shock/Stimulation		
									A26	Insufficient Device Problem Information		
5695	2018	All other reportable incidents	35852	Single-chamber implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A07	Electrical /Electronic Property Problem		
									A0709	Device Sensing Problem		
									A070909	Oversensing		
									A0713	Defibrillation/Stimulation Problem		
									A071302	Inappropriate/Inadequate Shock/Stimulation		
5696	2018	All other reportable incidents	35223	Endocardial/interventricular septal pacing lead	CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A05	Mechanical Problem	05/11/2018	
									A0510	Retraction Problem	05/11/2018	
5697	2017	TBD	47270	Cardiac resynchronization therapy implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	F1001	Absence of Treatment	A09	Output Problem		
									A0907	No Device Output		
5698	2018	All other reportable incidents	47270	Cardiac resynchronization therapy implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	F1905	Device Revision or Replacement	A07	Electrical /Electronic Property Problem		
									A0708	Power Problem		
5699	2018	All other reportable incidents	47270	Cardiac resynchronization therapy implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	F1905	Device Revision or Replacement	A07	Electrical /Electronic Property Problem	09/02/2018	
									A0708	Power Problem	09/02/2018	
5700	2017	All other reportable incidents			CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A07	Electrical /Electronic Property Problem		
									A0708	Power Problem		
5701	2018	All other reportable incidents	47270	Cardiac resynchronization therapy implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	F1905	Device Revision or Replacement	A07	Electrical /Electronic Property Problem	16/02/2018	
									A0708	Power Problem	16/02/2018	
5702	2018	Unanticipated serious deterioration in state of health			CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A07	Electrical /Electronic Property Problem	26/11/2018	
									A0708	Power Problem	26/11/2018	
									A13	Communication or Transmission Problem	26/11/2018	
									A26	Insufficient Device Problem Information	26/11/2018	
5703	2018	All other reportable incidents	47270	Cardiac resynchronization therapy implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	F1905	Device Revision or Replacement	A07	Electrical /Electronic Property Problem	04/07/2018	
									A0708	Power Problem	04/07/2018	
5704	2017	TBD	47263	Cardiac resynchronization therapy implantable pacemaker	CT1684	Cardiac pacemakers and associated devices	F1905	Device Revision or Replacement	A07	Electrical /Electronic Property Problem		
									A0708	Power Problem		
					CT1685	Cardiac pacemakers	F1905	Device Revision or Replacement	A07	Electrical /Electronic Property Problem		
									A0708	Power Problem		
					CT257	Implantable cardiac pacemakers	F1905	Device Revision or Replacement	A07	Electrical /Electronic Property Problem		
									A0708	Power Problem		
					CT2866	Implantable cardiac electrical stimulators	F1905	Device Revision or Replacement	A07	Electrical /Electronic Property Problem		
									A0708	Power Problem		
5705	2018	All other reportable incidents			CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A07	Electrical /Electronic Property Problem		
									A0709	Device Sensing Problem		
5706	2018	All other reportable incidents			CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A07	Electrical /Electronic Property Problem		
									A0709	Device Sensing Problem		
5707	2018	All other reportable incidents	47263	Cardiac resynchronization therapy implantable pacemaker	CT1684	Cardiac pacemakers and associated devices	F1905	Device Revision or Replacement	A07	Electrical /Electronic Property Problem	21/11/2018	
									A0708	Power Problem	21/11/2018	
					CT1685	Cardiac pacemakers	F1905	Device Revision or Replacement	A07	Electrical /Electronic Property Problem	21/11/2018	
									A0708	Power Problem	21/11/2018	
					CT257	Implantable cardiac pacemakers	F1905	Device Revision or Replacement	A07	Electrical /Electronic Property Problem	21/11/2018	
									A0708	Power Problem	21/11/2018	
					CT2866	Implantable cardiac electrical stimulators	F1905	Device Revision or Replacement	A07	Electrical /Electronic Property Problem	21/11/2018	
									A0708	Power Problem	21/11/2018	
5708	2018	TBD			CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A07	Electrical /Electronic Property Problem		
									A0709	Device Sensing Problem		
5709	2018	All other reportable incidents	35852	Single-chamber implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A07	Electrical /Electronic Property Problem		
									A0709	Device Sensing Problem		
5710	2017	TBD	35223	Endocardial/interventricular septal pacing lead	CT1684	Cardiac pacemakers and associated devices	E2401	Insufficient Information	A26	Insufficient Device Problem Information		
5711	2018	TBD	35852	Single-chamber implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A07	Electrical /Electronic Property Problem		
									A0709	Device Sensing Problem		
5712	2018	All other reportable incidents	35852	Single-chamber implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A07	Electrical /Electronic Property Problem		
									A0709	Device Sensing Problem		
5713	2019	All other reportable incidents	35223	Endocardial/interventricular septal pacing lead	CT1684	Cardiac pacemakers and associated devices	F1905	Device Revision or Replacement	A05	Mechanical Problem		
									A0722	Impedance Problem		
									A072201	High Impedance		
5714	2018	All other reportable incidents	35852	Single-chamber implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A07	Electrical /Electronic Property Problem		
									A0709	Device Sensing Problem		
5715	2018	All other reportable incidents			CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A05	Mechanical Problem	25/10/2018	
									A0506	Mechanical Jam	25/10/2018	
5716	2018	All other reportable incidents	47270	Cardiac resynchronization therapy implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	F1905	Device Revision or Replacement	A07	Electrical /Electronic Property Problem	23/07/2018	
									A0708	Power Problem	23/07/2018	
5717	2018	All other reportable incidents	35223	Endocardial/interventricular septal pacing lead	CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A04	Material Integrity Problem	25/10/2018	
									A0404	Crack	25/10/2018	
5718	2018	All other reportable incidents			CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A09	Output Problem		
									A0908	Incorrect, Inadequate or Imprecise Result or Readings		
5719	2017	All other reportable incidents			CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A09	Output Problem		
									A0907	No Device Output		
5720	2018	All other reportable incidents	35852	Single-chamber implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A07	Electrical /Electronic Property Problem		
									A0709	Device Sensing Problem		
5721	2017	All other reportable incidents			CT257	Implantable cardiac pacemakers	E2401	Insufficient Information	A26	Insufficient Device Problem Information		
5722	2017	TBD	47270	Cardiac resynchronization therapy implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A07	Electrical /Electronic Property Problem		
									A0708	Power Problem		
5723	2018	All other reportable incidents			CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A07	Electrical /Electronic Property Problem		
									A0709	Device Sensing Problem		
5724	2017	All other reportable incidents			CT1684	Cardiac pacemakers and associated devices	F05	Delay to Treatment/ Therapy	A09	Output Problem		
									A0908	Incorrect, Inadequate or Imprecise Result or Readings		

5725	2017	All other reportable incidents			CT257	Implantable cardiac pacemakers	F1905	Device Revision or Replacement	A07	Electrical /Electronic Property Problem	30/08/2017	
									A0708	Power Problem	30/08/2017	
5726	2018	All other reportable incidents			CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A09	Output Problem		
									A0907	No Device Output		
5727	2018	All other reportable incidents	47270	Cardiac resynchronization therapy implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A1304	Telemetry Discrepancy		
									A26	Insufficient Device Problem Information		
5728	2017	All other reportable incidents			CT257	Implantable cardiac pacemakers	F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A07	Electrical /Electronic Property Problem		
									A0709	Device Sensing Problem		
5729	2018	All other reportable incidents	47270	Cardiac resynchronization therapy implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	F1905	Device Revision or Replacement	A07	Electrical /Electronic Property Problem	05/07/2018	
									A0708	Power Problem	05/07/2018	
5730	2017	All other reportable incidents			CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A07	Electrical /Electronic Property Problem		
									A0708	Power Problem		
5731	2018	All other reportable incidents	47270	Cardiac resynchronization therapy implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	F1905	Device Revision or Replacement	A07	Electrical /Electronic Property Problem	10/11/2018	
									A0708	Power Problem	10/11/2018	
5732	2017	TBD	35852	Single-chamber implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A11	Computer Software Problem		
									A1102	Application Program Problem		
5733	2018	Unanticipated serious deterioration in state of health			CT1684	Cardiac pacemakers and associated devices	F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A09	Output Problem		
									A090404	Output Above Specifications		
									A15	Activation, Positioning or Separation Problem		
									A150103	Premature Activation		
5734	2018	All other reportable incidents	35223	Endocardial/interventricular septal pacing lead	CT1684	Cardiac pacemakers and associated devices	F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A07	Electrical /Electronic Property Problem	28/06/2018	
									A0709	Device Sensing Problem	28/06/2018	
									A15	Activation, Positioning or Separation Problem	28/06/2018	
									A150203	Difficult or Delayed Positioning	28/06/2018	
							F14	Prolonged Episode of Care	A07	Electrical /Electronic Property Problem	28/06/2018	
									A0709	Device Sensing Problem	28/06/2018	
									A15	Activation, Positioning or Separation Problem	28/06/2018	
									A150203	Difficult or Delayed Positioning	28/06/2018	
5735	2017	TBD			CT257	Implantable cardiac pacemakers	F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A07	Electrical /Electronic Property Problem		
									A0709	Device Sensing Problem		
5736	2017	All other reportable incidents	35223	Endocardial/interventricular septal pacing lead	CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A12	Connection Problem		
5737	2018	TBD	47270	Cardiac resynchronization therapy implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A07	Electrical /Electronic Property Problem		
									A0709	Device Sensing Problem		
5738	2018	TBD	47270	Cardiac resynchronization therapy implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A07	Electrical /Electronic Property Problem		
									A0708	Power Problem		
5739	2018	All other reportable incidents	35223	Endocardial/interventricular septal pacing lead	CT1684	Cardiac pacemakers and associated devices	F1905	Device Revision or Replacement	A07	Electrical /Electronic Property Problem		
									A0709	Device Sensing Problem		
5740	2018	TBD	35223	Endocardial/interventricular septal pacing lead	CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A15	Activation, Positioning or Separation Problem		
									A150103	Premature Activation		
5741	2017	TBD	35223	Endocardial/interventricular septal pacing lead	CT1684	Cardiac pacemakers and associated devices	F05	Delay to Treatment/ Therapy	A15	Activation, Positioning or Separation Problem		
									A150203	Difficult or Delayed Positioning		
5742	2017	All other reportable incidents	35852	Single-chamber implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A07	Electrical /Electronic Property Problem		
									A0709	Device Sensing Problem		
5743	2018	All other reportable incidents			CT257	Implantable cardiac pacemakers	F1905	Device Revision or Replacement	A07	Electrical /Electronic Property Problem	03/07/2018	
									A0708	Power Problem	03/07/2018	
5744	2018	All other reportable incidents	47270	Cardiac resynchronization therapy implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	F1905	Device Revision or Replacement	A07	Electrical /Electronic Property Problem	24/02/2018	
									A0708	Power Problem	24/02/2018	
5745	2018	All other reportable incidents			CT257	Implantable cardiac pacemakers	E0402	Hypersensitivity/Allergic Reaction	A24	No Apparent Device Problem		
									A26	Insufficient Device Problem Information		
5746	2018	All other reportable incidents			CT1684	Cardiac pacemakers and associated devices	E2401	Insufficient Information	A09	Output Problem		
									A0907	No Device Output		
5747	2018	All other reportable incidents	35852	Single-chamber implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A07	Electrical /Electronic Property Problem		
									A0709	Device Sensing Problem		
									A09	Output Problem		
									A090404	Output Above Specifications		
5748	2018	All other reportable incidents	47270	Cardiac resynchronization therapy implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A07	Electrical /Electronic Property Problem		
									A0709	Device Sensing Problem		
5749	2018	All other reportable incidents	47270	Cardiac resynchronization therapy implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	F1905	Device Revision or Replacement	A07	Electrical /Electronic Property Problem		
									A0708	Power Problem		
5750	2018	All other reportable incidents	35852	Single-chamber implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A07	Electrical /Electronic Property Problem		
									A0709	Device Sensing Problem		
5751	2018	All other reportable incidents	35223	Endocardial/interventricular septal pacing lead	CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A0104	Migration or Expulsion of Device	09/11/2018	
									A05	Mechanical Problem	09/11/2018	
									A0510	Retraction Problem	09/11/2018	
5752	2018	All other reportable incidents	35223	Endocardial/interventricular septal pacing lead	CT1684	Cardiac pacemakers and associated devices	F1905	Device Revision or Replacement	A05	Mechanical Problem	08/11/2018	
									A0510	Retraction Problem	08/11/2018	
5753	2018	All other reportable incidents			CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A09	Output Problem		
									A0907	No Device Output		
									A15	Activation, Positioning or Separation Problem		
									A150101	Activation Failure		
5754	2018	Unanticipated serious deterioration in state of health	35223	Endocardial/interventricular septal pacing lead	CT1684	Cardiac pacemakers and associated devices			A26	Insufficient Device Problem Information		
5755	2018	All other reportable incidents	47270	Cardiac resynchronization therapy implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A07	Electrical /Electronic Property Problem		
									A0722	Impedance Problem		
									A072201	High Impedance		
5756	2017	All other reportable incidents	47270	Cardiac resynchronization therapy implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A07	Electrical /Electronic Property Problem		
									A0709	Device Sensing Problem		
5757	2018	All other reportable incidents			CT257	Implantable cardiac pacemakers	F14	Prolonged Episode of Care	A13	Communication or Transmission Problem		
									A26	Insufficient Device Problem Information		
5758	2017	TBD			CT257	Implantable cardiac pacemakers	F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A07	Electrical /Electronic Property Problem		
									A0709	Device Sensing Problem		
5759	2018	All other reportable incidents			CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A07	Electrical /Electronic Property Problem		
									A0708	Power Problem		
5760	2017	All other reportable incidents			CT257	Implantable cardiac pacemakers	E2401	Insufficient Information	A07	Electrical /Electronic Property Problem		
									A0708	Power Problem		
5761	2017	TBD			CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A07	Electrical /Electronic Property Problem		
									A0709	Device Sensing Problem		

5762	2017	TBD	35223	Endocardial/interventricular septal pacing lead	CT1684	Cardiac pacemakers and associated devices			A26	Insufficient Device Problem Information		
5763	2017	All other reportable incidents	35852	Single-chamber implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	F05	Delay to Treatment/ Therapy	A07	Electrical /Electronic Property Problem		
									A0709	Device Sensing Problem		
5764	2017	All other reportable incidents	47270	Cardiac resynchronization therapy implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A07	Electrical /Electronic Property Problem		
									A0709	Device Sensing Problem		
5765	2018	TBD	35852	Single-chamber implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A07	Electrical /Electronic Property Problem		
									A0709	Device Sensing Problem		
5766	2018	TBD	35223	Endocardial/interventricular septal pacing lead	CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A15	Activation, Positioning or Separation Problem	17/01/2018	
									A150203	Difficult or Delayed Positioning	17/01/2018	
5767	2018	All other reportable incidents	35852	Single-chamber implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A0104	Migration or Expulsion of Device		
5768	2018	TBD			CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A07	Electrical /Electronic Property Problem		
									A0709	Device Sensing Problem		
5769	2018	All other reportable incidents			CT257	Implantable cardiac pacemakers	F1905	Device Revision or Replacement	A07	Electrical /Electronic Property Problem	26/11/2018	
									A0708	Power Problem	26/11/2018	
									A13	Communication or Transmission Problem	26/11/2018	
									A26	Insufficient Device Problem Information	26/11/2018	
5770	2018	All other reportable incidents	35852	Single-chamber implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A07	Electrical /Electronic Property Problem		
									A0709	Device Sensing Problem		
5771	2017	All other reportable incidents			CT257	Implantable cardiac pacemakers	F1905	Device Revision or Replacement	A13	Communication or Transmission Problem		
									A26	Insufficient Device Problem Information		
5772	2018	All other reportable incidents	35852	Single-chamber implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A07	Electrical /Electronic Property Problem	13/11/2018	
									A0709	Device Sensing Problem	13/11/2018	
5773	2018	All other reportable incidents	35852	Single-chamber implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A07	Electrical /Electronic Property Problem		
									A0709	Device Sensing Problem		
5774	2018	TBD	47270	Cardiac resynchronization therapy implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	F1905	Device Revision or Replacement	A12	Connection Problem	05/02/2018	
5775	2018	All other reportable incidents	35223	Endocardial/interventricular septal pacing lead	CT1684	Cardiac pacemakers and associated devices	E2401	Insufficient Information	A07	Electrical /Electronic Property Problem		
									A0709	Device Sensing Problem		
5776	2018	All other reportable incidents	47270	Cardiac resynchronization therapy implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	F1905	Device Revision or Replacement	A07	Electrical /Electronic Property Problem	23/10/2018	
									A0708	Power Problem	23/10/2018	
5777	2017	TBD	35223	Endocardial/interventricular septal pacing lead	CT1684	Cardiac pacemakers and associated devices	F14	Prolonged Episode of Care	A26	Insufficient Device Problem Information		
5778	2018	TBD	47263	Cardiac resynchronization therapy implantable pacemaker	CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A07	Electrical /Electronic Property Problem		
									A0709	Device Sensing Problem		
					CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A07	Electrical /Electronic Property Problem		
									A0709	Device Sensing Problem		
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A07	Electrical /Electronic Property Problem		
									A0709	Device Sensing Problem		
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A07	Electrical /Electronic Property Problem		
									A0709	Device Sensing Problem		
5779	2018	All other reportable incidents	47263	Cardiac resynchronization therapy implantable pacemaker	CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A0104	Migration or Expulsion of Device		
									A07	Electrical /Electronic Property Problem		
									A0709	Device Sensing Problem		
					CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A0104	Migration or Expulsion of Device		
									A07	Electrical /Electronic Property Problem		
									A0709	Device Sensing Problem		
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A0104	Migration or Expulsion of Device		
									A07	Electrical /Electronic Property Problem		
									A0709	Device Sensing Problem		
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A0104	Migration or Expulsion of Device		
									A07	Electrical /Electronic Property Problem		
									A0709	Device Sensing Problem		
5780	2017	TBD	35223	Endocardial/interventricular septal pacing lead	CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A15	Activation, Positioning or Separation Problem		
									A150203	Difficult or Delayed Positioning		
5781	2018	All other reportable incidents	35852	Single-chamber implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	F1905	Device Revision or Replacement	A07	Electrical /Electronic Property Problem	27/11/2018	
									A0708	Power Problem	27/11/2018	
5782	2018	All other reportable incidents	35223	Endocardial/interventricular septal pacing lead	CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A07	Electrical /Electronic Property Problem		
									A0709	Device Sensing Problem		
5783	2017	All other reportable incidents	35852	Single-chamber implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A07	Electrical /Electronic Property Problem		
									A0709	Device Sensing Problem		
									A09	Output Problem		
									A090404	Output Above Specifications		
5784	2018	All other reportable incidents	35039	Epicardial pacing lead	CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A07	Electrical /Electronic Property Problem		
									A0709	Device Sensing Problem		
5785	2017	All other reportable incidents			CT257	Implantable cardiac pacemakers	F1905	Device Revision or Replacement	A1104	Computer Operating System Problem		
5786	2018	All other reportable incidents			CT1684	Cardiac pacemakers and associated devices	F05	Delay to Treatment/ Therapy	A09	Output Problem		
									A0907	No Device Output		
5787	2018	All other reportable incidents			CT1684	Cardiac pacemakers and associated devices	F05	Delay to Treatment/ Therapy	A09	Output Problem		
									A0907	No Device Output		
5788	2018	All other reportable incidents			CT257	Implantable cardiac pacemakers	F1905	Device Revision or Replacement	A07	Electrical /Electronic Property Problem	15/11/2018	
									A0708	Power Problem	15/11/2018	
5789	2017	All other reportable incidents	47270	Cardiac resynchronization therapy implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A07	Electrical /Electronic Property Problem		
									A0709	Device Sensing Problem		
5790	2018	All other reportable incidents	35852	Single-chamber implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	F1905	Device Revision or Replacement	A07	Electrical /Electronic Property Problem	07/05/2018	
									A0709	Device Sensing Problem	07/05/2018	
5791	2017	All other reportable incidents	47270	Cardiac resynchronization therapy implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A07	Electrical /Electronic Property Problem		
									A0708	Power Problem		
5792	2017	TBD	47270	Cardiac resynchronization therapy implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A07	Electrical /Electronic Property Problem		
									A0709	Device Sensing Problem		
							F14	Prolonged Episode of Care	A07	Electrical /Electronic Property Problem		
									A0709	Device Sensing Problem		
5793	2018	TBD	35223	Endocardial/interventricular septal pacing lead	CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A15	Activation, Positioning or Separation Problem		
									A150301	Separation Failure		
5794	2018	All other reportable incidents			CT257	Implantable cardiac pacemakers	F1905	Device Revision or Replacement	A07	Electrical /Electronic Property Problem	14/11/2018	
									A0708	Power Problem	14/11/2018	
5795	2018	All other reportable incidents			CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A09	Output Problem		
									A0907	No Device Output		

5796	2018	All other reportable incidents			CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A15	Activation, Positioning or Separation Problem		
									A150101	Activation Failure		
5797	2018	All other reportable incidents			CT1684	Cardiac pacemakers and associated devices	F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A07	Electrical /Electronic Property Problem		
									A0708	Power Problem		
5798	2018	All other reportable incidents			CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A11	Computer Software Problem		
									A1102	Application Program Problem		
5799	2018	All other reportable incidents	35852	Single-chamber implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A07	Electrical /Electronic Property Problem		
									A0709	Device Sensing Problem		
5800	2018	All other reportable incidents	47263	Cardiac resynchronization therapy implantable pacemaker	CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A07	Electrical /Electronic Property Problem		
									A0709	Device Sensing Problem		
					CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A07	Electrical /Electronic Property Problem		
									A0709	Device Sensing Problem		
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A07	Electrical /Electronic Property Problem		
									A0709	Device Sensing Problem		
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A07	Electrical /Electronic Property Problem		
									A0709	Device Sensing Problem		
5801	2018	TBD	35852	Single-chamber implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A09	Output Problem		
									A090404	Output Above Specifications		
5802	2018	All other reportable incidents	35223	Endocardial/interventricular septal pacing lead	CT1684	Cardiac pacemakers and associated devices	F1905	Device Revision or Replacement	A07	Electrical /Electronic Property Problem	17/07/2014	
									A0709	Device Sensing Problem	17/07/2014	
5803	2018	All other reportable incidents	35223	Endocardial/interventricular septal pacing lead	CT1684	Cardiac pacemakers and associated devices	F1905	Device Revision or Replacement	A07	Electrical /Electronic Property Problem	07/11/2013	
									A0709	Device Sensing Problem	07/11/2013	
5804	2018	All other reportable incidents	35223	Endocardial/interventricular septal pacing lead	CT1684	Cardiac pacemakers and associated devices	F1905	Device Revision or Replacement	A07	Electrical /Electronic Property Problem	04/01/2017	
									A0709	Device Sensing Problem	04/01/2017	
5805	2018	All other reportable incidents	35223	Endocardial/interventricular septal pacing lead	CT1684	Cardiac pacemakers and associated devices	F1905	Device Revision or Replacement	A07	Electrical /Electronic Property Problem	14/04/2016	
									A0709	Device Sensing Problem	14/04/2016	
5806	2018	All other reportable incidents	35223	Endocardial/interventricular septal pacing lead	CT1684	Cardiac pacemakers and associated devices	F1905	Device Revision or Replacement	A07	Electrical /Electronic Property Problem	26/10/2018	
									A0709	Device Sensing Problem	26/10/2018	
5807	2018	TBD	35223	Endocardial/interventricular septal pacing lead	CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A07	Electrical /Electronic Property Problem		
									A0709	Device Sensing Problem		
5808	2018	All other reportable incidents			CT257	Implantable cardiac pacemakers	F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A07	Electrical /Electronic Property Problem		
									A0709	Device Sensing Problem		
5809	2018	All other reportable incidents	35223	Endocardial/interventricular septal pacing lead	CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A07	Electrical /Electronic Property Problem		
									A0709	Device Sensing Problem		
5810	2018	All other reportable incidents	47270	Cardiac resynchronization therapy implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	F1905	Device Revision or Replacement	A07	Electrical /Electronic Property Problem	08/11/2018	
									A0709	Device Sensing Problem	08/11/2018	
5811	2017	All other reportable incidents	35852	Single-chamber implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A09	Output Problem		
									A0907	No Device Output		
5812	2018	All other reportable incidents	35223	Endocardial/interventricular septal pacing lead	CT1684	Cardiac pacemakers and associated devices	F1905	Device Revision or Replacement	A0104	Migration or Expulsion of Device		
									A07	Electrical /Electronic Property Problem		
									A0709	Device Sensing Problem		
5813	2018	All other reportable incidents			CT257	Implantable cardiac pacemakers	F1905	Device Revision or Replacement	A07	Electrical /Electronic Property Problem	15/11/2018	
									A0708	Power Problem	15/11/2018	
5814	2018	All other reportable incidents	47270	Cardiac resynchronization therapy implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A11	Computer Software Problem		
									A1102	Application Program Problem		
5815	2018	All other reportable incidents	47270	Cardiac resynchronization therapy implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A26	Insufficient Device Problem Information		
5816	2018	All other reportable incidents			CT257	Implantable cardiac pacemakers	F1905	Device Revision or Replacement	A07	Electrical /Electronic Property Problem		
									A0708	Power Problem		
5817	2017	TBD			CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A07	Electrical /Electronic Property Problem		
									A0708	Power Problem		
5818	2018	All other reportable incidents	35852	Single-chamber implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A07	Electrical /Electronic Property Problem		
									A0709	Device Sensing Problem		
5819	2017	All other reportable incidents			CT257	Implantable cardiac pacemakers	F1905	Device Revision or Replacement	A07	Electrical /Electronic Property Problem		
									A0709	Device Sensing Problem		
5820	2017	TBD	47270	Cardiac resynchronization therapy implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A07	Electrical /Electronic Property Problem		
									A0708	Power Problem		
5821	2017	Death			CT257	Implantable cardiac pacemakers	F21	Unexpected Deterioration	A09	Output Problem	02/06/2017	
									A0907	No Device Output	02/06/2017	
5822	2017	Unanticipated serious deterioration in state of health	47270	Cardiac resynchronization therapy implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E2401	Insufficient Information	A26	Insufficient Device Problem Information		
5823	2017	TBD			CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A11	Computer Software Problem		
									A1102	Application Program Problem		
5824	2017	TBD	47270	Cardiac resynchronization therapy implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A07	Electrical /Electronic Property Problem		
									A0709	Device Sensing Problem		
5825	2017	TBD	35223	Endocardial/interventricular septal pacing lead	CT1684	Cardiac pacemakers and associated devices			A26	Insufficient Device Problem Information		
5826	2018	TBD	35852	Single-chamber implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A07	Electrical /Electronic Property Problem		
									A0709	Device Sensing Problem		
5827	2018	All other reportable incidents	35223	Endocardial/interventricular septal pacing lead	CT1684	Cardiac pacemakers and associated devices	F1905	Device Revision or Replacement	A0104	Migration or Expulsion of Device		
5828	2018	TBD	35223	Endocardial/interventricular septal pacing lead	CT1684	Cardiac pacemakers and associated devices	F1905	Device Revision or Replacement	A07	Electrical /Electronic Property Problem	25/02/2018	
									A0709	Device Sensing Problem	25/02/2018	
5829	2018	TBD	35223	Endocardial/interventricular septal pacing lead	CT1684	Cardiac pacemakers and associated devices			A26	Insufficient Device Problem Information		
5830	2018	All other reportable incidents	35223	Endocardial/interventricular septal pacing lead	CT1684	Cardiac pacemakers and associated devices	F1905	Device Revision or Replacement	A05	Mechanical Problem	27/07/2018	
									A0510	Retraction Problem	27/07/2018	
5831	2018	All other reportable incidents			CT257	Implantable cardiac pacemakers	F1905	Device Revision or Replacement	A07	Electrical /Electronic Property Problem		
									A0708	Power Problem		
5832	2018	All other reportable incidents	47263	Cardiac resynchronization therapy implantable pacemaker	CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A07	Electrical /Electronic Property Problem		
									A0708	Power Problem		
					CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A07	Electrical /Electronic Property Problem		
									A0708	Power Problem		
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A07	Electrical /Electronic Property Problem		
									A0708	Power Problem		
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A07	Electrical /Electronic Property Problem		
									A0708	Power Problem		
5833	2018	All other reportable incidents	35223	Endocardial/interventricular septal pacing lead	CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A0104	Migration or Expulsion of Device		

5834	2018	All other reportable incidents	47270	Cardiac resynchronization therapy implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A07	Electrical /Electronic Property Problem		
									A0708	Power Problem		
5835	2018	All other reportable incidents			CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A07	Electrical /Electronic Property Problem		
									A0709	Device Sensing Problem		
5836	2018	All other reportable incidents			CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A13	Communication or Transmission Problem	23/07/2018	
									A26	Insufficient Device Problem Information	23/07/2018	
5837	2018	All other reportable incidents	47263	Cardiac resynchronization therapy implantable pacemaker	CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A07	Electrical /Electronic Property Problem		
									A0709	Device Sensing Problem		
					CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A07	Electrical /Electronic Property Problem		
									A0709	Device Sensing Problem		
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A07	Electrical /Electronic Property Problem		
									A0709	Device Sensing Problem		
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A07	Electrical /Electronic Property Problem		
									A0709	Device Sensing Problem		
5838	2018	All other reportable incidents			CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A16	Protective Measures Problem		
									A1602	Fail-Safe Problem		
5839	2018	All other reportable incidents			CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A07	Electrical /Electronic Property Problem		
									A0708	Power Problem		
5840	2018	All other reportable incidents	35223	Endocardial/interventricular septal pacing lead	CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A07	Electrical /Electronic Property Problem	19/11/2018	
									A0709	Device Sensing Problem	19/11/2018	
									A15	Activation, Positioning or Separation Problem	19/11/2018	
									A150101	Activation Failure	19/11/2018	
5841	2018	All other reportable incidents			CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A07	Electrical /Electronic Property Problem		
									A0708	Power Problem		
5842	2018	All other reportable incidents			CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A07	Electrical /Electronic Property Problem	07/12/2018	
									A0708	Power Problem	07/12/2018	
5843	2018	All other reportable incidents	47270	Cardiac resynchronization therapy implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A07	Electrical /Electronic Property Problem		
									A0709	Device Sensing Problem		
5844	2018	All other reportable incidents			CT257	Implantable cardiac pacemakers	F1905	Device Revision or Replacement	A07	Electrical /Electronic Property Problem	27/11/2018	
									A0709	Device Sensing Problem	27/11/2018	
5845	2018	Unanticipated serious deterioration in state of health			CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A07	Electrical /Electronic Property Problem		
									A0708	Power Problem		
5846	2018	All other reportable incidents			CT257	Implantable cardiac pacemakers	F1905	Device Revision or Replacement	A07	Electrical /Electronic Property Problem	06/07/2018	
									A0708	Power Problem	06/07/2018	
5847	2018	All other reportable incidents			CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A07	Electrical /Electronic Property Problem	22/06/2018	
									A0708	Power Problem	22/06/2018	
5848	2017	TBD			CT257	Implantable cardiac pacemakers	F1905	Device Revision or Replacement	A07	Electrical /Electronic Property Problem		
									A0709	Device Sensing Problem		
5849	2018	TBD	47270	Cardiac resynchronization therapy implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A07	Electrical /Electronic Property Problem		
									A0709	Device Sensing Problem		
5850	2017	All other reportable incidents	35223	Endocardial/interventricular septal pacing lead	CT1684	Cardiac pacemakers and associated devices	F1905	Device Revision or Replacement	A07	Electrical /Electronic Property Problem		
									A0709	Device Sensing Problem		
5851	2018	All other reportable incidents	47270	Cardiac resynchronization therapy implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	F1905	Device Revision or Replacement	A07	Electrical /Electronic Property Problem	20/02/2018	
									A0708	Power Problem	20/02/2018	
5852	2018	All other reportable incidents	47270	Cardiac resynchronization therapy implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	F1905	Device Revision or Replacement	A07	Electrical /Electronic Property Problem	22/02/2018	
									A0708	Power Problem	22/02/2018	
5853	2017	TBD			CT257	Implantable cardiac pacemakers			A26	Insufficient Device Problem Information		
5854	2018	All other reportable incidents			CT257	Implantable cardiac pacemakers	F1905	Device Revision or Replacement	A07	Electrical /Electronic Property Problem	02/10/2018	
									A0708	Power Problem	02/10/2018	
5855	2017	TBD	35223	Endocardial/interventricular septal pacing lead	CT1684	Cardiac pacemakers and associated devices	F14	Prolonged Episode of Care	A0104	Migration or Expulsion of Device		
5856	2017	Unanticipated serious deterioration in state of health	35852	Single-chamber implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A07	Electrical /Electronic Property Problem		
									A0709	Device Sensing Problem		
5857	2018	All other reportable incidents	37265	Dual-chamber implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E2401	Insufficient Information	A07	Electrical /Electronic Property Problem	20/09/2022	20/09/2022
									A0709	Device Sensing Problem	20/09/2022	20/09/2022
									A070909	Oversensing	20/09/2022	20/09/2022
							E2403	No Clinical Signs, Symptoms or Conditions	A07	Electrical /Electronic Property Problem	20/09/2022	20/09/2022
									A0709	Device Sensing Problem	20/09/2022	20/09/2022
									A070909	Oversensing	20/09/2022	20/09/2022
							F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A07	Electrical /Electronic Property Problem	20/09/2022	20/09/2022
									A0709	Device Sensing Problem	20/09/2022	20/09/2022
									A070909	Oversensing	20/09/2022	20/09/2022
							F15	Recognized Device or Procedural Complication	A07	Electrical /Electronic Property Problem	20/09/2022	20/09/2022
									A0709	Device Sensing Problem	20/09/2022	20/09/2022
									A070909	Oversensing	20/09/2022	20/09/2022
5858	2018	All other reportable incidents			CT257	Implantable cardiac pacemakers	F1905	Device Revision or Replacement	A07	Electrical /Electronic Property Problem	04/09/2018	
									A0708	Power Problem	04/09/2018	
5859	2017	All other reportable incidents			CT257	Implantable cardiac pacemakers	F1905	Device Revision or Replacement	A07	Electrical /Electronic Property Problem	20/09/2017	
									A0708	Power Problem	20/09/2017	
5860	2018	All other reportable incidents	47270	Cardiac resynchronization therapy implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A07	Electrical /Electronic Property Problem		
									A0708	Power Problem		
5861	2018	All other reportable incidents			CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A12	Connection Problem	09/10/2018	
									A1204	Failure to Disconnect	09/10/2018	
5862	2018	TBD	47270	Cardiac resynchronization therapy implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A07	Electrical /Electronic Property Problem		
									A0709	Device Sensing Problem		
5863	2018	All other reportable incidents			CT257	Implantable cardiac pacemakers	F1905	Device Revision or Replacement	A07	Electrical /Electronic Property Problem		
									A0708	Power Problem		
5864	2018	All other reportable incidents			CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A07	Electrical /Electronic Property Problem		
									A0709	Device Sensing Problem		
5865	2018	TBD	35852	Single-chamber implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A0203	Defective Device		
									A26	Insufficient Device Problem Information		
5866	2017	All other reportable incidents			CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A04	Material Integrity Problem		
									A0404	Crack		
									A10	Temperature Problem		
									A1005	Overheating of Device		
5867	2017	Unanticipated serious deterioration in state of health	35223	Endocardial/interventricular septal pacing lead	CT1684	Cardiac pacemakers and associated devices	F1905	Device Revision or Replacement	A0104	Migration or Expulsion of Device	05/05/2017	

5868	2018	Unanticipated serious deterioration in state of health			CT257	Implantable cardiac pacemakers	E2401	Insufficient Information	A26	Insufficient Device Problem Information		
5869	2017	TBD	35223	Endocardial/interventricular septal pacing lead	CT1684	Cardiac pacemakers and associated devices			A26	Insufficient Device Problem Information		
5870	2018	All other reportable incidents	47270	Cardiac resynchronization therapy implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	F1905	Device Revision or Replacement	A07	Electrical /Electronic Property Problem	29/11/2018	
									A0708	Power Problem	29/11/2018	
5871	2018	All other reportable incidents	47270	Cardiac resynchronization therapy implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	F1905	Device Revision or Replacement	A07	Electrical /Electronic Property Problem		
									A0708	Power Problem		
5872	2017	TBD	35223	Endocardial/interventricular septal pacing lead	CT1684	Cardiac pacemakers and associated devices	E2401	Insufficient Information	A26	Insufficient Device Problem Information		
5873	2018	All other reportable incidents	35223	Endocardial/interventricular septal pacing lead	CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A05	Mechanical Problem		
									A0510	Retraction Problem		
5874	2017	TBD			CT257	Implantable cardiac pacemakers	F1905	Device Revision or Replacement	A07	Electrical /Electronic Property Problem		
									A0709	Device Sensing Problem		
									A09	Output Problem		
									A0907	No Device Output		
5875	2018	All other reportable incidents	35223	Endocardial/interventricular septal pacing lead	CT1684	Cardiac pacemakers and associated devices	F14	Prolonged Episode of Care	A05	Mechanical Problem	07/12/2018	
									A0510	Retraction Problem	07/12/2018	
							F1905	Device Revision or Replacement	A05	Mechanical Problem	07/12/2018	
									A0510	Retraction Problem	07/12/2018	
5876	2017	All other reportable incidents			CT257	Implantable cardiac pacemakers	F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A07	Electrical /Electronic Property Problem		
									A0709	Device Sensing Problem		
5877	2018	All other reportable incidents	35223	Endocardial/interventricular septal pacing lead	CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A15	Activation, Positioning or Separation Problem	03/12/2018	
									A150101	Activation Failure	03/12/2018	
5878	2018	All other reportable incidents	35223	Endocardial/interventricular septal pacing lead	CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A05	Mechanical Problem	26/11/2018	
									A0510	Retraction Problem	26/11/2018	
5879	2017	TBD	35223	Endocardial/interventricular septal pacing lead	CT1684	Cardiac pacemakers and associated devices			A26	Insufficient Device Problem Information		
5880	2017	TBD	35223	Endocardial/interventricular septal pacing lead	CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A15	Activation, Positioning or Separation Problem		
									A150101	Activation Failure		
5881	2018	All other reportable incidents	47270	Cardiac resynchronization therapy implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A07	Electrical /Electronic Property Problem		
									A0708	Power Problem		
									A070802	Intermittent Loss of Power		
5882	2017	Unanticipated serious deterioration in state of health			CT257	Implantable cardiac pacemakers	F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A09	Output Problem	24/05/2017	
									A0907	No Device Output	24/05/2017	
							F1905	Device Revision or Replacement	A09	Output Problem	24/05/2017	
									A0907	No Device Output	24/05/2017	
5883	2018	All other reportable incidents	35223	Endocardial/interventricular septal pacing lead	CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A07	Electrical /Electronic Property Problem		
									A0709	Device Sensing Problem		
5884	2018	All other reportable incidents			CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A26	Insufficient Device Problem Information		
5885	2018	All other reportable incidents	60789	Intracardiac pacemaker	CT130	Catheters and associated devices	E2401	Insufficient Information	A07	Electrical /Electronic Property Problem	03/06/2019	
									A0708	Power Problem	03/06/2019	
									A0709	Device Sensing Problem	03/06/2019	
					CT145	Cardiovascular devices	E2401	Insufficient Information	A07	Electrical /Electronic Property Problem	03/06/2019	
									A0708	Power Problem	03/06/2019	
									A0709	Device Sensing Problem	03/06/2019	
					CT1684	Cardiac pacemakers and associated devices	E2401	Insufficient Information	A07	Electrical /Electronic Property Problem	03/06/2019	
									A0708	Power Problem	03/06/2019	
									A0709	Device Sensing Problem	03/06/2019	
					CT1685	Cardiac pacemakers	E2401	Insufficient Information	A07	Electrical /Electronic Property Problem	03/06/2019	
									A0708	Power Problem	03/06/2019	
									A0709	Device Sensing Problem	03/06/2019	
					CT233	Surgical	E2401	Insufficient Information	A07	Electrical /Electronic Property Problem	03/06/2019	
									A0708	Power Problem	03/06/2019	
									A0709	Device Sensing Problem	03/06/2019	
					CT2406	Implantable	E2401	Insufficient Information	A07	Electrical /Electronic Property Problem	03/06/2019	
									A0708	Power Problem	03/06/2019	
									A0709	Device Sensing Problem	03/06/2019	
					CT2407	Active implantable	E2401	Insufficient Information	A07	Electrical /Electronic Property Problem	03/06/2019	
									A0708	Power Problem	03/06/2019	
									A0709	Device Sensing Problem	03/06/2019	
					CT257	Implantable cardiac pacemakers	E2401	Insufficient Information	A07	Electrical /Electronic Property Problem	03/06/2019	
									A0708	Power Problem	03/06/2019	
									A0709	Device Sensing Problem	03/06/2019	
					CT2727	Active implant devices/ancillaries	E2401	Insufficient Information	A07	Electrical /Electronic Property Problem	03/06/2019	
									A0708	Power Problem	03/06/2019	
									A0709	Device Sensing Problem	03/06/2019	
					CT2750	Implantable cardiac electrical stimulators and associated devices	E2401	Insufficient Information	A07	Electrical /Electronic Property Problem	03/06/2019	
									A0708	Power Problem	03/06/2019	
									A0709	Device Sensing Problem	03/06/2019	
					CT2818	Catheterization/Cannulation/Fluid path devices	E2401	Insufficient Information	A07	Electrical /Electronic Property Problem	03/06/2019	
									A0708	Power Problem	03/06/2019	
									A0709	Device Sensing Problem	03/06/2019	
					CT2866	Implantable cardiac electrical stimulators	E2401	Insufficient Information	A07	Electrical /Electronic Property Problem	03/06/2019	
									A0708	Power Problem	03/06/2019	
									A0709	Device Sensing Problem	03/06/2019	
					CT2985	Active	E2401	Insufficient Information	A07	Electrical /Electronic Property Problem	03/06/2019	
									A0708	Power Problem	03/06/2019	
									A0709	Device Sensing Problem	03/06/2019	
					CT3001	Implantable metal	E2401	Insufficient Information	A07	Electrical /Electronic Property Problem	03/06/2019	
									A0708	Power Problem	03/06/2019	
									A0709	Device Sensing Problem	03/06/2019	
					CT3002	Implantable synthetic polymer	E2401	Insufficient Information	A07	Electrical /Electronic Property Problem	03/06/2019	
									A0708	Power Problem	03/06/2019	
									A0709	Device Sensing Problem	03/06/2019	
					CT3006	Implantable inorganic material	E2401	Insufficient Information	A07	Electrical /Electronic Property Problem	03/06/2019	
									A0708	Power Problem	03/06/2019	
									A0709	Device Sensing Problem	03/06/2019	
					CT314	Electrically powered	E2401	Insufficient Information	A07	Electrical /Electronic Property Problem	03/06/2019	

									A0708	Power Problem	03/06/2019	
									A0709	Device Sensing Problem	03/06/2019	
					CT321	Long-term surgical invasive	E2401	Insufficient Information	A07	Electrical /Electronic Property Problem	03/06/2019	
									A0708	Power Problem	03/06/2019	
									A0709	Device Sensing Problem	03/06/2019	
					CT981	Single use	E2401	Insufficient Information	A07	Electrical /Electronic Property Problem	03/06/2019	
									A0708	Power Problem	03/06/2019	
									A0709	Device Sensing Problem	03/06/2019	
					CT983	Surgical invasive	E2401	Insufficient Information	A07	Electrical /Electronic Property Problem	03/06/2019	
									A0708	Power Problem	03/06/2019	
									A0709	Device Sensing Problem	03/06/2019	
5886	2018	All other reportable incidents	35852	Single-chamber implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A07	Electrical /Electronic Property Problem		
									A0709	Device Sensing Problem		
5887	2017	All other reportable incidents	35223	Endocardial/interventricular septal pacing lead	CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A07	Electrical /Electronic Property Problem		
									A0708	Power Problem		
5888	2018	All other reportable incidents	35852	Single-chamber implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A07	Electrical /Electronic Property Problem		
									A0709	Device Sensing Problem		
5889	2018	All other reportable incidents	35852	Single-chamber implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A07	Electrical /Electronic Property Problem		
									A0709	Device Sensing Problem		
5890	2017	All other reportable incidents	47270	Cardiac resynchronization therapy implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A07	Electrical /Electronic Property Problem	18/05/2017	
									A0708	Power Problem	18/05/2017	
5891	2017	All other reportable incidents	35223	Endocardial/interventricular septal pacing lead	CT1684	Cardiac pacemakers and associated devices	F1905	Device Revision or Replacement	A07	Electrical /Electronic Property Problem		
									A0709	Device Sensing Problem		
5892	2018	All other reportable incidents			CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A09	Output Problem		
									A0907	No Device Output		
5893	2018	All other reportable incidents			CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A07	Electrical /Electronic Property Problem	06/09/2018	
									A0708	Power Problem	06/09/2018	
5894	2018	All other reportable incidents	47270	Cardiac resynchronization therapy implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E1906	Unspecified Infection	A01	Patient Device Interaction Problem	22/02/2018	
									A17	Compatibility Problem	22/02/2018	
							F1905	Device Revision or Replacement	A01	Patient Device Interaction Problem	22/02/2018	
									A17	Compatibility Problem	22/02/2018	
5895	2018	All other reportable incidents	47270	Cardiac resynchronization therapy implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	F1905	Device Revision or Replacement	A07	Electrical /Electronic Property Problem		
									A0722	Impedance Problem		
									A072201	High Impedance		
5896	2018	All other reportable incidents	35223	Endocardial/interventricular septal pacing lead	CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A05	Mechanical Problem	13/12/2018	
									A0510	Retraction Problem	13/12/2018	
5897	2018	Death	35852	Single-chamber implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E2401	Insufficient Information	A09	Output Problem		
									A090404	Output Above Specifications		
									A26	Insufficient Device Problem Information		
							F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A09	Output Problem		
									A090404	Output Above Specifications		
									A26	Insufficient Device Problem Information		
5898	2017	All other reportable incidents	35223	Endocardial/interventricular septal pacing lead	CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A12	Connection Problem		
5899	2018	All other reportable incidents	47270	Cardiac resynchronization therapy implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	F1905	Device Revision or Replacement	A07	Electrical /Electronic Property Problem	07/08/2018	
									A0708	Power Problem	07/08/2018	
5900	2018	All other reportable incidents			CT257	Implantable cardiac pacemakers	F1905	Device Revision or Replacement	A0709	Device Sensing Problem	24/11/2018	
									A070909	Oversensing	24/11/2018	
									A09	Output Problem	24/11/2018	
									A0904	Energy Output Problem	24/11/2018	
									A090402	Failure to Deliver Energy	24/11/2018	
5901	2017	All other reportable incidents	35223	Endocardial/interventricular septal pacing lead	CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A0104	Migration or Expulsion of Device		
									A07	Electrical /Electronic Property Problem		
									A0709	Device Sensing Problem		
5902	2017	All other reportable incidents	35223	Endocardial/interventricular septal pacing lead	CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A05	Mechanical Problem		
									A0510	Retraction Problem		
									A07	Electrical /Electronic Property Problem		
									A0709	Device Sensing Problem		
									A15	Activation, Positioning or Separation Problem		
									A150203	Difficult or Delayed Positioning		
5903	2018	All other reportable incidents			CT257	Implantable cardiac pacemakers	F1905	Device Revision or Replacement	A07	Electrical /Electronic Property Problem		
									A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
5904	2018	All other reportable incidents			CT257	Implantable cardiac pacemakers	F1905	Device Revision or Replacement	A1304	Telemetry Discrepancy	24/07/2018	
									A26	Insufficient Device Problem Information	24/07/2018	
5905	2018	All other reportable incidents	47270	Cardiac resynchronization therapy implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A07	Electrical /Electronic Property Problem	11/07/2018	
									A0708	Power Problem	11/07/2018	
5906	2018	All other reportable incidents			CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A09	Output Problem		
									A0907	No Device Output		
5907	2017	All other reportable incidents	47270	Cardiac resynchronization therapy implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A07	Electrical /Electronic Property Problem		
									A0709	Device Sensing Problem		
5908	2017	TBD	35223	Endocardial/interventricular septal pacing lead	CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A15	Activation, Positioning or Separation Problem		
									A150203	Difficult or Delayed Positioning		
5909	2017	All other reportable incidents	35223	Endocardial/interventricular septal pacing lead	CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A0104	Migration or Expulsion of Device		
									A05	Mechanical Problem		
									A0510	Retraction Problem		
5910	2017	All other reportable incidents	35223	Endocardial/interventricular septal pacing lead	CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A15	Activation, Positioning or Separation Problem		
									A150101	Activation Failure		
5911	2018	All other reportable incidents	35852	Single-chamber implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	F14	Prolonged Episode of Care	A09	Output Problem		
									A0904	Energy Output Problem		
5912	2018	All other reportable incidents	35852	Single-chamber implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	F14	Prolonged Episode of Care	A05	Mechanical Problem		
									A0510	Retraction Problem		
5913	2018	All other reportable incidents	47270	Cardiac resynchronization therapy implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	F1905	Device Revision or Replacement	A07	Electrical /Electronic Property Problem		
									A0708	Power Problem		
5914	2018	All other reportable incidents	35852	Single-chamber implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A07	Electrical /Electronic Property Problem		
									A0709	Device Sensing Problem		
									A070909	Oversensing		
5915	2018	All other reportable incidents			CT257	Implantable cardiac pacemakers	F1905	Device Revision or Replacement	A07	Electrical /Electronic Property Problem		

									A0708	Power Problem		
5916	2018	All other reportable incidents	47270	Cardiac resynchronization therapy implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	F1905	Device Revision or Replacement	A07	Electrical /Electronic Property Problem	14/06/2018	
									A0709	Device Sensing Problem	14/06/2018	
5917	2017	All other reportable incidents	47270	Cardiac resynchronization therapy implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	F1905	Device Revision or Replacement	A07	Electrical /Electronic Property Problem	15/05/2017	
									A0708	Power Problem	15/05/2017	
5918	2017	All other reportable incidents	47270	Cardiac resynchronization therapy implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	F1905	Device Revision or Replacement	A07	Electrical /Electronic Property Problem	23/05/2017	
									A0708	Power Problem	23/05/2017	
5919	2018	All other reportable incidents	47270	Cardiac resynchronization therapy implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A07	Electrical /Electronic Property Problem		
									A0709	Device Sensing Problem		
5920	2017	All other reportable incidents	47270	Cardiac resynchronization therapy implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A07	Electrical /Electronic Property Problem		
									A0709	Device Sensing Problem		
5921	2017	All other reportable incidents			CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A07	Electrical /Electronic Property Problem		
									A0708	Power Problem		
5922	2017	TBD			CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A0203	Defective Device		
									A26	Insufficient Device Problem Information		
5923	2018	All other reportable incidents	47270	Cardiac resynchronization therapy implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	F1905	Device Revision or Replacement	A07	Electrical /Electronic Property Problem	24/12/2018	
									A0705	Battery Problem	24/12/2018	
									A070504	Premature Discharge of Battery	24/12/2018	
5924	2018	All other reportable incidents	35852	Single-chamber implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A07	Electrical /Electronic Property Problem		
									A0713	Defibrillation/Stimulation Problem		
									A071302	Inappropriate/Inadequate Shock/Stimulation		
5925	2018	All other reportable incidents			CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A05	Mechanical Problem		
									A0506	Mechanical Jam		
5926	2017	TBD	47270	Cardiac resynchronization therapy implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A07	Electrical /Electronic Property Problem		
									A0708	Power Problem		
5927	2018	All other reportable incidents	47263	Cardiac resynchronization therapy implantable pacemaker	CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A07	Electrical /Electronic Property Problem		
									A0709	Device Sensing Problem		
					CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A07	Electrical /Electronic Property Problem		
									A0709	Device Sensing Problem		
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A07	Electrical /Electronic Property Problem		
									A0709	Device Sensing Problem		
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A07	Electrical /Electronic Property Problem		
									A0709	Device Sensing Problem		
5928	2017	TBD	35223	Endocardial/interventricular septal pacing lead	CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A0104	Migration or Expulsion of Device		
5929	2017	All other reportable incidents	35223	Endocardial/interventricular septal pacing lead	CT1684	Cardiac pacemakers and associated devices	F1905	Device Revision or Replacement	A07	Electrical /Electronic Property Problem		
									A0709	Device Sensing Problem		
									A15	Activation, Positioning or Separation Problem		
									A150203	Difficult or Delayed Positioning		
5930	2018	All other reportable incidents	47270	Cardiac resynchronization therapy implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A07	Electrical /Electronic Property Problem		
									A0709	Device Sensing Problem		
5931	2017	All other reportable incidents	47270	Cardiac resynchronization therapy implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A07	Electrical /Electronic Property Problem		
									A0708	Power Problem		
5932	2017	TBD	35852	Single-chamber implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A07	Electrical /Electronic Property Problem		
									A0709	Device Sensing Problem		
5933	2018	All other reportable incidents	35223	Endocardial/interventricular septal pacing lead	CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A05	Mechanical Problem	30/11/2018	
									A0510	Retraction Problem	30/11/2018	
5934	2018	All other reportable incidents	35852	Single-chamber implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	F1905	Device Revision or Replacement	A07	Electrical /Electronic Property Problem	09/08/2018	
									A0708	Power Problem	09/08/2018	
5935	2017	TBD			CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A07	Electrical /Electronic Property Problem		
									A0708	Power Problem		
5936	2017	All other reportable incidents	47270	Cardiac resynchronization therapy implantable defibrillator	CT2866	Implantable cardiac electrical stimulators			A07	Electrical /Electronic Property Problem		
									A0708	Power Problem		
5937	2018	All other reportable incidents			CT257	Implantable cardiac pacemakers	F1905	Device Revision or Replacement	A07	Electrical /Electronic Property Problem		
									A0708	Power Problem		
5938	2018	All other reportable incidents	47270	Cardiac resynchronization therapy implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A07	Electrical /Electronic Property Problem		
									A0708	Power Problem		
5939	2018	All other reportable incidents			CT1684	Cardiac pacemakers and associated devices	E2401	Insufficient Information	A09	Output Problem		
									A0907	No Device Output		
5940	2017	All other reportable incidents	47270	Cardiac resynchronization therapy implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A07	Electrical /Electronic Property Problem		
									A0709	Device Sensing Problem		
5941	2017	All other reportable incidents	35223	Endocardial/interventricular septal pacing lead	CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A07	Electrical /Electronic Property Problem		
									A0709	Device Sensing Problem		
5942	2018	All other reportable incidents	35223	Endocardial/interventricular septal pacing lead	CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A04	Material Integrity Problem		
									A0410	Material Perforation		
5943	2018	All other reportable incidents	35852	Single-chamber implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A07	Electrical /Electronic Property Problem		
									A0709	Device Sensing Problem		
									A09	Output Problem		
									A090404	Output Above Specifications		
5944	2017	TBD			CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A07	Electrical /Electronic Property Problem		
									A0709	Device Sensing Problem		
5945	2018	All other reportable incidents			CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A07	Electrical /Electronic Property Problem		
									A0708	Power Problem		
5946	2017	All other reportable incidents	47263	Cardiac resynchronization therapy implantable pacemaker	CT1684	Cardiac pacemakers and associated devices	F1905	Device Revision or Replacement	A07	Electrical /Electronic Property Problem	06/02/2017	
									A0708	Power Problem	06/02/2017	
					CT1685	Cardiac pacemakers	F1905	Device Revision or Replacement	A07	Electrical /Electronic Property Problem	06/02/2017	
									A0708	Power Problem	06/02/2017	
					CT257	Implantable cardiac pacemakers	F1905	Device Revision or Replacement	A07	Electrical /Electronic Property Problem	06/02/2017	
									A0708	Power Problem	06/02/2017	
					CT2866	Implantable cardiac electrical stimulators	F1905	Device Revision or Replacement	A07	Electrical /Electronic Property Problem	06/02/2017	
									A0708	Power Problem	06/02/2017	
5947	2017	TBD	35223	Endocardial/interventricular septal pacing lead	CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A04	Material Integrity Problem		
									A0413	Material Separation		
5948	2017	All other reportable incidents	35223	Endocardial/interventricular septal pacing lead	CT1684	Cardiac pacemakers and associated devices	F14	Prolonged Episode of Care	A15	Activation, Positioning or Separation Problem	11/10/2017	
									A150101	Activation Failure	11/10/2017	
5949	2017	TBD	35223	Endocardial/interventricular septal pacing lead	CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem		
									A26	Insufficient Device Problem Information		
5950	2017	TBD	47270	Cardiac resynchronization therapy implantable	CT2866	Implantable cardiac electrical stimulators	F1905	Device Revision or Replacement	A07	Electrical /Electronic Property Problem		

				defibrillator					A0708	Power Problem		
5951	2017	All other reportable incidents			CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A07	Electrical /Electronic Property Problem		
									A0709	Device Sensing Problem		
5952	2017	All other reportable incidents			CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A07	Electrical /Electronic Property Problem		
									A0708	Power Problem		
5953	2017	TBD	47270	Cardiac resynchronization therapy implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A07	Electrical /Electronic Property Problem		
									A0708	Power Problem		
5954	2017	All other reportable incidents	35223	Endocardial/interventricular septal pacing lead	CT1684	Cardiac pacemakers and associated devices	E2401	Insufficient Information	A26	Insufficient Device Problem Information		
5955	2017	TBD	35223	Endocardial/interventricular septal pacing lead	CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A15	Activation, Positioning or Separation Problem		
									A150101	Activation Failure		
5956	2017	All other reportable incidents	35852	Single-chamber implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A07	Electrical /Electronic Property Problem		
									A0709	Device Sensing Problem		
5957	2017	TBD	47270	Cardiac resynchronization therapy implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	F1905	Device Revision or Replacement	A07	Electrical /Electronic Property Problem		
									A0708	Power Problem		
5958	2017	All other reportable incidents			CT257	Implantable cardiac pacemakers	F1905	Device Revision or Replacement	A07	Electrical /Electronic Property Problem	16/10/2017	
									A0708	Power Problem	16/10/2017	
5959	2017	TBD			CT257	Implantable cardiac pacemakers	F1905	Device Revision or Replacement	A07	Electrical /Electronic Property Problem		
									A0708	Power Problem		
5960	2017	All other reportable incidents	35852	Single-chamber implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A07	Electrical /Electronic Property Problem		
									A0709	Device Sensing Problem		
5961	2017	All other reportable incidents	35223	Endocardial/interventricular septal pacing lead	CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A07	Electrical /Electronic Property Problem		
									A0709	Device Sensing Problem		
5962	2017	TBD	47270	Cardiac resynchronization therapy implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A07	Electrical /Electronic Property Problem		
									A0709	Device Sensing Problem		
									A09	Output Problem		
									A090404	Output Above Specifications		
5963	2021	All other reportable incidents	17882	Physiologic-monitoring defibrillation system	CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem		
							F27	Problem Identified During Non-Clinical Procedure	A0705	Battery Problem		
5964	2020	All other reportable incidents	36241	Pacemaker/defibrillator lead	CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070909	Oversensing		
									A090801	Signal Artifact/Noise		
							F24	Insufficient Health Impact Information	A070909	Oversensing		
									A090801	Signal Artifact/Noise		
5965	2020	All other reportable incidents	17882	Physiologic-monitoring defibrillation system	CT1684	Cardiac pacemakers and associated devices	E060101	Asystole	A070908	Failure to Sense		
							F21	Unexpected Deterioration	A070908	Failure to Sense		
							F26	No Health Consequences or Impact	A070908	Failure to Sense		
5966	2022	All other reportable incidents	17882	Physiologic-monitoring defibrillation system	CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A0501	Detachment of Device or Device Component		
							F27	Problem Identified During Non-Clinical Procedure	A0501	Detachment of Device or Device Component		
5967	2022	All other reportable incidents	47263	Cardiac resynchronization therapy implantable pacemaker	CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A0908	Incorrect, Inadequate or Imprecise Result or Readings	15/02/2022	15/02/2022
							F08	Hospitalization or Prolonged Hospitalization	A0908	Incorrect, Inadequate or Imprecise Result or Readings	15/02/2022	15/02/2022
							F1903	Device Explantation	A0908	Incorrect, Inadequate or Imprecise Result or Readings	15/02/2022	15/02/2022
					CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A0908	Incorrect, Inadequate or Imprecise Result or Readings	15/02/2022	15/02/2022
							F08	Hospitalization or Prolonged Hospitalization	A0908	Incorrect, Inadequate or Imprecise Result or Readings	15/02/2022	15/02/2022
							F1903	Device Explantation	A0908	Incorrect, Inadequate or Imprecise Result or Readings	15/02/2022	15/02/2022
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A0908	Incorrect, Inadequate or Imprecise Result or Readings	15/02/2022	15/02/2022
							F08	Hospitalization or Prolonged Hospitalization	A0908	Incorrect, Inadequate or Imprecise Result or Readings	15/02/2022	15/02/2022
							F1903	Device Explantation	A0908	Incorrect, Inadequate or Imprecise Result or Readings	15/02/2022	15/02/2022
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A0908	Incorrect, Inadequate or Imprecise Result or Readings	15/02/2022	15/02/2022
							F08	Hospitalization or Prolonged Hospitalization	A0908	Incorrect, Inadequate or Imprecise Result or Readings	15/02/2022	15/02/2022
							F1903	Device Explantation	A0908	Incorrect, Inadequate or Imprecise Result or Readings	15/02/2022	15/02/2022
5968	2022	All other reportable incidents	47263	Cardiac resynchronization therapy implantable pacemaker	CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A0908	Incorrect, Inadequate or Imprecise Result or Readings		
							F26	No Health Consequences or Impact	A0908	Incorrect, Inadequate or Imprecise Result or Readings		
					CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A0908	Incorrect, Inadequate or Imprecise Result or Readings		
							F26	No Health Consequences or Impact	A0908	Incorrect, Inadequate or Imprecise Result or Readings		
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A0908	Incorrect, Inadequate or Imprecise Result or Readings		
							F26	No Health Consequences or Impact	A0908	Incorrect, Inadequate or Imprecise Result or Readings		
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A0908	Incorrect, Inadequate or Imprecise Result or Readings		
							F26	No Health Consequences or Impact	A0908	Incorrect, Inadequate or Imprecise Result or Readings		
5969	2020	All other reportable incidents	47265	Dual-chamber implantable pacemaker, rate-responsive	CT141	Electrodes/Leads and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A0203	Defective Device	02/10/2020	02/10/2020
									A070909	Oversensing	02/10/2020	02/10/2020
							F1901	Additional Surgery	A0203	Defective Device	02/10/2020	02/10/2020
									A070909	Oversensing	02/10/2020	02/10/2020
							F1905	Device Revision or Replacement	A0203	Defective Device	02/10/2020	02/10/2020
									A070909	Oversensing	02/10/2020	02/10/2020
					CT145	Cardiovascular devices	E2403	No Clinical Signs, Symptoms or Conditions	A0203	Defective Device	02/10/2020	02/10/2020
									A070909	Oversensing	02/10/2020	02/10/2020
							F1901	Additional Surgery	A0203	Defective Device	02/10/2020	02/10/2020
									A070909	Oversensing	02/10/2020	02/10/2020
							F1905	Device Revision or Replacement	A0203	Defective Device	02/10/2020	02/10/2020
									A070909	Oversensing	02/10/2020	02/10/2020
					CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A0203	Defective Device	02/10/2020	02/10/2020
									A070909	Oversensing	02/10/2020	02/10/2020
							F1901	Additional Surgery	A0203	Defective Device	02/10/2020	02/10/2020

							A070909	Oversensing	02/10/2020	02/10/2020
						F1905	Device Revision or Replacement	A0203	Defective Device	02/10/2020	02/10/2020
								A070909	Oversensing	02/10/2020	02/10/2020
					CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A0203	Defective Device	02/10/2020	02/10/2020
									A070909	Oversensing	02/10/2020	02/10/2020
							F1901	Additional Surgery	A0203	Defective Device	02/10/2020	02/10/2020
									A070909	Oversensing	02/10/2020	02/10/2020
							F1905	Device Revision or Replacement	A0203	Defective Device	02/10/2020	02/10/2020
									A070909	Oversensing	02/10/2020	02/10/2020
					CT233	Surgical	E2403	No Clinical Signs, Symptoms or Conditions	A0203	Defective Device	02/10/2020	02/10/2020
									A070909	Oversensing	02/10/2020	02/10/2020
							F1901	Additional Surgery	A0203	Defective Device	02/10/2020	02/10/2020
									A070909	Oversensing	02/10/2020	02/10/2020
							F1905	Device Revision or Replacement	A0203	Defective Device	02/10/2020	02/10/2020
									A070909	Oversensing	02/10/2020	02/10/2020
					CT2406	Implantable	E2403	No Clinical Signs, Symptoms or Conditions	A0203	Defective Device	02/10/2020	02/10/2020
									A070909	Oversensing	02/10/2020	02/10/2020
							F1901	Additional Surgery	A0203	Defective Device	02/10/2020	02/10/2020
									A070909	Oversensing	02/10/2020	02/10/2020
							F1905	Device Revision or Replacement	A0203	Defective Device	02/10/2020	02/10/2020
									A070909	Oversensing	02/10/2020	02/10/2020
					CT2407	Active implantable	E2403	No Clinical Signs, Symptoms or Conditions	A0203	Defective Device	02/10/2020	02/10/2020
									A070909	Oversensing	02/10/2020	02/10/2020
							F1901	Additional Surgery	A0203	Defective Device	02/10/2020	02/10/2020
									A070909	Oversensing	02/10/2020	02/10/2020
							F1905	Device Revision or Replacement	A0203	Defective Device	02/10/2020	02/10/2020
									A070909	Oversensing	02/10/2020	02/10/2020
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A0203	Defective Device	02/10/2020	02/10/2020
									A070909	Oversensing	02/10/2020	02/10/2020
							F1901	Additional Surgery	A0203	Defective Device	02/10/2020	02/10/2020
									A070909	Oversensing	02/10/2020	02/10/2020
							F1905	Device Revision or Replacement	A0203	Defective Device	02/10/2020	02/10/2020
									A070909	Oversensing	02/10/2020	02/10/2020
					CT2727	Active implant devices/ancillaries	E2403	No Clinical Signs, Symptoms or Conditions	A0203	Defective Device	02/10/2020	02/10/2020
									A070909	Oversensing	02/10/2020	02/10/2020
							F1901	Additional Surgery	A0203	Defective Device	02/10/2020	02/10/2020
									A070909	Oversensing	02/10/2020	02/10/2020
							F1905	Device Revision or Replacement	A0203	Defective Device	02/10/2020	02/10/2020
									A070909	Oversensing	02/10/2020	02/10/2020
					CT2750	Implantable cardiac electrical stimulators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A0203	Defective Device	02/10/2020	02/10/2020
									A070909	Oversensing	02/10/2020	02/10/2020
							F1901	Additional Surgery	A0203	Defective Device	02/10/2020	02/10/2020
									A070909	Oversensing	02/10/2020	02/10/2020
							F1905	Device Revision or Replacement	A0203	Defective Device	02/10/2020	02/10/2020
									A070909	Oversensing	02/10/2020	02/10/2020
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A0203	Defective Device	02/10/2020	02/10/2020
									A070909	Oversensing	02/10/2020	02/10/2020
							F1901	Additional Surgery	A0203	Defective Device	02/10/2020	02/10/2020
									A070909	Oversensing	02/10/2020	02/10/2020
							F1905	Device Revision or Replacement	A0203	Defective Device	02/10/2020	02/10/2020
									A070909	Oversensing	02/10/2020	02/10/2020
					CT2985	Active	E2403	No Clinical Signs, Symptoms or Conditions	A0203	Defective Device	02/10/2020	02/10/2020
									A070909	Oversensing	02/10/2020	02/10/2020
							F1901	Additional Surgery	A0203	Defective Device	02/10/2020	02/10/2020
									A070909	Oversensing	02/10/2020	02/10/2020
							F1905	Device Revision or Replacement	A0203	Defective Device	02/10/2020	02/10/2020
									A070909	Oversensing	02/10/2020	02/10/2020
					CT3001	Implantable metal	E2403	No Clinical Signs, Symptoms or Conditions	A0203	Defective Device	02/10/2020	02/10/2020
									A070909	Oversensing	02/10/2020	02/10/2020
							F1901	Additional Surgery	A0203	Defective Device	02/10/2020	02/10/2020
									A070909	Oversensing	02/10/2020	02/10/2020
							F1905	Device Revision or Replacement	A0203	Defective Device	02/10/2020	02/10/2020
									A070909	Oversensing	02/10/2020	02/10/2020
					CT3002	Implantable synthetic polymer	E2403	No Clinical Signs, Symptoms or Conditions	A0203	Defective Device	02/10/2020	02/10/2020
									A070909	Oversensing	02/10/2020	02/10/2020
							F1901	Additional Surgery	A0203	Defective Device	02/10/2020	02/10/2020
									A070909	Oversensing	02/10/2020	02/10/2020
							F1905	Device Revision or Replacement	A0203	Defective Device	02/10/2020	02/10/2020
									A070909	Oversensing	02/10/2020	02/10/2020
					CT3006	Implantable inorganic material	E2403	No Clinical Signs, Symptoms or Conditions	A0203	Defective Device	02/10/2020	02/10/2020
									A070909	Oversensing	02/10/2020	02/10/2020
							F1901	Additional Surgery	A0203	Defective Device	02/10/2020	02/10/2020
									A070909	Oversensing	02/10/2020	02/10/2020
							F1905	Device Revision or Replacement	A0203	Defective Device	02/10/2020	02/10/2020
									A070909	Oversensing	02/10/2020	02/10/2020
					CT314	Electrically powered	E2403	No Clinical Signs, Symptoms or Conditions	A0203	Defective Device	02/10/2020	02/10/2020
									A070909	Oversensing	02/10/2020	02/10/2020
							F1901	Additional Surgery	A0203	Defective Device	02/10/2020	02/10/2020
									A070909	Oversensing	02/10/2020	02/10/2020
							F1905	Device Revision or Replacement	A0203	Defective Device	02/10/2020	02/10/2020
									A070909	Oversensing	02/10/2020	02/10/2020
					CT321	Long-term surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A0203	Defective Device	02/10/2020	02/10/2020
									A070909	Oversensing	02/10/2020	02/10/2020
							F1901	Additional Surgery	A0203	Defective Device	02/10/2020	02/10/2020
									A070909	Oversensing	02/10/2020	02/10/2020
							F1905	Device Revision or Replacement	A0203	Defective Device	02/10/2020	02/10/2020
									A070909	Oversensing	02/10/2020	02/10/2020
					CT449	Generators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A0203	Defective Device	02/10/2020	02/10/2020

									A070909	Oversensing	02/10/2020	02/10/2020
							F1901	Additional Surgery	A0203	Defective Device	02/10/2020	02/10/2020
									A070909	Oversensing	02/10/2020	02/10/2020
							F1905	Device Revision or Replacement	A0203	Defective Device	02/10/2020	02/10/2020
									A070909	Oversensing	02/10/2020	02/10/2020
					CT981	Single use	E2403	No Clinical Signs, Symptoms or Conditions	A0203	Defective Device	02/10/2020	02/10/2020
									A070909	Oversensing	02/10/2020	02/10/2020
							F1901	Additional Surgery	A0203	Defective Device	02/10/2020	02/10/2020
									A070909	Oversensing	02/10/2020	02/10/2020
							F1905	Device Revision or Replacement	A0203	Defective Device	02/10/2020	02/10/2020
									A070909	Oversensing	02/10/2020	02/10/2020
					CT983	Surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A0203	Defective Device	02/10/2020	02/10/2020
									A070909	Oversensing	02/10/2020	02/10/2020
							F1901	Additional Surgery	A0203	Defective Device	02/10/2020	02/10/2020
									A070909	Oversensing	02/10/2020	02/10/2020
							F1905	Device Revision or Replacement	A0203	Defective Device	02/10/2020	02/10/2020
									A070909	Oversensing	02/10/2020	02/10/2020
5970	2022	All other reportable incidents	47265	Dual-chamber implantable pacemaker, rate-responsive	CT141	Electrodes/Leads and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	27/03/2022	27/03/2022
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	27/03/2022	27/03/2022
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	27/03/2022	27/03/2022
					CT145	Cardiovascular devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	27/03/2022	27/03/2022
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	27/03/2022	27/03/2022
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	27/03/2022	27/03/2022
					CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	27/03/2022	27/03/2022
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	27/03/2022	27/03/2022
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	27/03/2022	27/03/2022
					CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	27/03/2022	27/03/2022
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	27/03/2022	27/03/2022
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	27/03/2022	27/03/2022
					CT233	Surgical	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	27/03/2022	27/03/2022
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	27/03/2022	27/03/2022
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	27/03/2022	27/03/2022
					CT2406	Implantable	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	27/03/2022	27/03/2022
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	27/03/2022	27/03/2022
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	27/03/2022	27/03/2022
					CT2407	Active implantable	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	27/03/2022	27/03/2022
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	27/03/2022	27/03/2022
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	27/03/2022	27/03/2022
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	27/03/2022	27/03/2022
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	27/03/2022	27/03/2022
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	27/03/2022	27/03/2022
					CT2727	Active implant devices/ancillaries	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	27/03/2022	27/03/2022
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	27/03/2022	27/03/2022
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	27/03/2022	27/03/2022
					CT2750	Implantable cardiac electrical stimulators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	27/03/2022	27/03/2022
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	27/03/2022	27/03/2022
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	27/03/2022	27/03/2022
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	27/03/2022	27/03/2022
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	27/03/2022	27/03/2022
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	27/03/2022	27/03/2022
					CT2985	Active	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	27/03/2022	27/03/2022
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	27/03/2022	27/03/2022
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	27/03/2022	27/03/2022
					CT3001	Implantable metal	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	27/03/2022	27/03/2022
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	27/03/2022	27/03/2022
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	27/03/2022	27/03/2022
					CT3002	Implantable synthetic polymer	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	27/03/2022	27/03/2022
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	27/03/2022	27/03/2022
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	27/03/2022	27/03/2022
					CT3006	Implantable inorganic material	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	27/03/2022	27/03/2022
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	27/03/2022	27/03/2022
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	27/03/2022	27/03/2022
					CT314	Electrically powered	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	27/03/2022	27/03/2022
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	27/03/2022	27/03/2022
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	27/03/2022	27/03/2022
					CT321	Long-term surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	27/03/2022	27/03/2022
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	27/03/2022	27/03/2022
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	27/03/2022	27/03/2022
					CT449	Generators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	27/03/2022	27/03/2022
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	27/03/2022	27/03/2022
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	27/03/2022	27/03/2022
					CT981	Single use	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	27/03/2022	27/03/2022
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	27/03/2022	27/03/2022
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	27/03/2022	27/03/2022
					CT983	Surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	27/03/2022	27/03/2022
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	27/03/2022	27/03/2022
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	27/03/2022	27/03/2022
5971	2020	All other reportable incidents	47265	Dual-chamber implantable pacemaker, rate-responsive	CT141	Electrodes/Leads and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070909	Oversensing	09/10/2020	09/10/2020
							F1901	Additional Surgery	A070909	Oversensing	09/10/2020	09/10/2020
							F1905	Device Revision or Replacement	A070909	Oversensing	09/10/2020	09/10/2020
					CT145	Cardiovascular devices	E2403	No Clinical Signs, Symptoms or Conditions	A070909	Oversensing	09/10/2020	09/10/2020
							F1901	Additional Surgery	A070909	Oversensing	09/10/2020	09/10/2020
							F1905	Device Revision or Replacement	A070909	Oversensing	09/10/2020	09/10/2020
					CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070909	Oversensing	09/10/2020	09/10/2020
							F1901	Additional Surgery	A070909	Oversensing	09/10/2020	09/10/2020
							F1905	Device Revision or Replacement	A070909	Oversensing	09/10/2020	09/10/2020
					CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A070909	Oversensing	09/10/2020	09/10/2020
							F1901	Additional Surgery	A070909	Oversensing	09/10/2020	09/10/2020

							F1905	Device Revision or Replacement	A070909	Oversensing	09/10/2020	09/10/2020
					CT233	Surgical	E2403	No Clinical Signs, Symptoms or Conditions	A070909	Oversensing	09/10/2020	09/10/2020
							F1901	Additional Surgery	A070909	Oversensing	09/10/2020	09/10/2020
							F1905	Device Revision or Replacement	A070909	Oversensing	09/10/2020	09/10/2020
					CT2406	Implantable	E2403	No Clinical Signs, Symptoms or Conditions	A070909	Oversensing	09/10/2020	09/10/2020
							F1901	Additional Surgery	A070909	Oversensing	09/10/2020	09/10/2020
							F1905	Device Revision or Replacement	A070909	Oversensing	09/10/2020	09/10/2020
					CT2407	Active implantable	E2403	No Clinical Signs, Symptoms or Conditions	A070909	Oversensing	09/10/2020	09/10/2020
							F1901	Additional Surgery	A070909	Oversensing	09/10/2020	09/10/2020
							F1905	Device Revision or Replacement	A070909	Oversensing	09/10/2020	09/10/2020
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A070909	Oversensing	09/10/2020	09/10/2020
							F1901	Additional Surgery	A070909	Oversensing	09/10/2020	09/10/2020
							F1905	Device Revision or Replacement	A070909	Oversensing	09/10/2020	09/10/2020
					CT2727	Active implant devices/ancillaries	E2403	No Clinical Signs, Symptoms or Conditions	A070909	Oversensing	09/10/2020	09/10/2020
							F1901	Additional Surgery	A070909	Oversensing	09/10/2020	09/10/2020
							F1905	Device Revision or Replacement	A070909	Oversensing	09/10/2020	09/10/2020
					CT2750	Implantable cardiac electrical stimulators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070909	Oversensing	09/10/2020	09/10/2020
							F1901	Additional Surgery	A070909	Oversensing	09/10/2020	09/10/2020
							F1905	Device Revision or Replacement	A070909	Oversensing	09/10/2020	09/10/2020
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A070909	Oversensing	09/10/2020	09/10/2020
							F1901	Additional Surgery	A070909	Oversensing	09/10/2020	09/10/2020
							F1905	Device Revision or Replacement	A070909	Oversensing	09/10/2020	09/10/2020
					CT2985	Active	E2403	No Clinical Signs, Symptoms or Conditions	A070909	Oversensing	09/10/2020	09/10/2020
							F1901	Additional Surgery	A070909	Oversensing	09/10/2020	09/10/2020
							F1905	Device Revision or Replacement	A070909	Oversensing	09/10/2020	09/10/2020
					CT3001	Implantable metal	E2403	No Clinical Signs, Symptoms or Conditions	A070909	Oversensing	09/10/2020	09/10/2020
							F1901	Additional Surgery	A070909	Oversensing	09/10/2020	09/10/2020
							F1905	Device Revision or Replacement	A070909	Oversensing	09/10/2020	09/10/2020
					CT3002	Implantable synthetic polymer	E2403	No Clinical Signs, Symptoms or Conditions	A070909	Oversensing	09/10/2020	09/10/2020
							F1901	Additional Surgery	A070909	Oversensing	09/10/2020	09/10/2020
							F1905	Device Revision or Replacement	A070909	Oversensing	09/10/2020	09/10/2020
					CT3006	Implantable inorganic material	E2403	No Clinical Signs, Symptoms or Conditions	A070909	Oversensing	09/10/2020	09/10/2020
							F1901	Additional Surgery	A070909	Oversensing	09/10/2020	09/10/2020
							F1905	Device Revision or Replacement	A070909	Oversensing	09/10/2020	09/10/2020
					CT314	Electrically powered	E2403	No Clinical Signs, Symptoms or Conditions	A070909	Oversensing	09/10/2020	09/10/2020
							F1901	Additional Surgery	A070909	Oversensing	09/10/2020	09/10/2020
							F1905	Device Revision or Replacement	A070909	Oversensing	09/10/2020	09/10/2020
					CT321	Long-term surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A070909	Oversensing	09/10/2020	09/10/2020
							F1901	Additional Surgery	A070909	Oversensing	09/10/2020	09/10/2020
							F1905	Device Revision or Replacement	A070909	Oversensing	09/10/2020	09/10/2020
					CT449	Generators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070909	Oversensing	09/10/2020	09/10/2020
							F1901	Additional Surgery	A070909	Oversensing	09/10/2020	09/10/2020
							F1905	Device Revision or Replacement	A070909	Oversensing	09/10/2020	09/10/2020
					CT981	Single use	E2403	No Clinical Signs, Symptoms or Conditions	A070909	Oversensing	09/10/2020	09/10/2020
							F1901	Additional Surgery	A070909	Oversensing	09/10/2020	09/10/2020
							F1905	Device Revision or Replacement	A070909	Oversensing	09/10/2020	09/10/2020
					CT983	Surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A070909	Oversensing	09/10/2020	09/10/2020
							F1901	Additional Surgery	A070909	Oversensing	09/10/2020	09/10/2020
							F1905	Device Revision or Replacement	A070909	Oversensing	09/10/2020	09/10/2020
5972	2020	All other reportable incidents	47267	Single-chamber implantable pacemaker, rate-responsive	CT141	Electrodes/Leads and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	16/11/2020	16/11/2020
							F12	Serious Injury/ Illness/ Impairment	A070504	Premature Discharge of Battery	16/11/2020	16/11/2020
							F19	Surgical Intervention	A070504	Premature Discharge of Battery	16/11/2020	16/11/2020
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	16/11/2020	16/11/2020
					CT145	Cardiovascular devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	16/11/2020	16/11/2020
							F12	Serious Injury/ Illness/ Impairment	A070504	Premature Discharge of Battery	16/11/2020	16/11/2020
							F19	Surgical Intervention	A070504	Premature Discharge of Battery	16/11/2020	16/11/2020
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	16/11/2020	16/11/2020
					CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	16/11/2020	16/11/2020
							F12	Serious Injury/ Illness/ Impairment	A070504	Premature Discharge of Battery	16/11/2020	16/11/2020
							F19	Surgical Intervention	A070504	Premature Discharge of Battery	16/11/2020	16/11/2020
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	16/11/2020	16/11/2020
					CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	16/11/2020	16/11/2020
							F12	Serious Injury/ Illness/ Impairment	A070504	Premature Discharge of Battery	16/11/2020	16/11/2020
							F19	Surgical Intervention	A070504	Premature Discharge of Battery	16/11/2020	16/11/2020
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	16/11/2020	16/11/2020
					CT233	Surgical	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	16/11/2020	16/11/2020
							F12	Serious Injury/ Illness/ Impairment	A070504	Premature Discharge of Battery	16/11/2020	16/11/2020
							F19	Surgical Intervention	A070504	Premature Discharge of Battery	16/11/2020	16/11/2020
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	16/11/2020	16/11/2020
					CT2406	Implantable	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	16/11/2020	16/11/2020
							F12	Serious Injury/ Illness/ Impairment	A070504	Premature Discharge of Battery	16/11/2020	16/11/2020
							F19	Surgical Intervention	A070504	Premature Discharge of Battery	16/11/2020	16/11/2020
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	16/11/2020	16/11/2020
					CT2407	Active implantable	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	16/11/2020	16/11/2020
							F12	Serious Injury/ Illness/ Impairment	A070504	Premature Discharge of Battery	16/11/2020	16/11/2020
							F19	Surgical Intervention	A070504	Premature Discharge of Battery	16/11/2020	16/11/2020
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	16/11/2020	16/11/2020
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	16/11/2020	16/11/2020
							F12	Serious Injury/ Illness/ Impairment	A070504	Premature Discharge of Battery	16/11/2020	16/11/2020
							F19	Surgical Intervention	A070504	Premature Discharge of Battery	16/11/2020	16/11/2020
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	16/11/2020	16/11/2020
					CT2727	Active implant devices/ancillaries	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	16/11/2020	16/11/2020
							F12	Serious Injury/ Illness/ Impairment	A070504	Premature Discharge of Battery	16/11/2020	16/11/2020
							F19	Surgical Intervention	A070504	Premature Discharge of Battery	16/11/2020	16/11/2020
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	16/11/2020	16/11/2020
					CT2750	Implantable cardiac electrical stimulators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	16/11/2020	16/11/2020
							F12	Serious Injury/ Illness/ Impairment	A070504	Premature Discharge of Battery	16/11/2020	16/11/2020
							F19	Surgical Intervention	A070504	Premature Discharge of Battery	16/11/2020	16/11/2020

							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	16/11/2020	16/11/2020
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	16/11/2020	16/11/2020
							F12	Serious Injury/ Illness/ Impairment	A070504	Premature Discharge of Battery	16/11/2020	16/11/2020
							F19	Surgical Intervention	A070504	Premature Discharge of Battery	16/11/2020	16/11/2020
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	16/11/2020	16/11/2020
					CT2985	Active	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	16/11/2020	16/11/2020
							F12	Serious Injury/ Illness/ Impairment	A070504	Premature Discharge of Battery	16/11/2020	16/11/2020
							F19	Surgical Intervention	A070504	Premature Discharge of Battery	16/11/2020	16/11/2020
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	16/11/2020	16/11/2020
					CT3001	Implantable metal	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	16/11/2020	16/11/2020
							F12	Serious Injury/ Illness/ Impairment	A070504	Premature Discharge of Battery	16/11/2020	16/11/2020
							F19	Surgical Intervention	A070504	Premature Discharge of Battery	16/11/2020	16/11/2020
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	16/11/2020	16/11/2020
					CT3002	Implantable synthetic polymer	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	16/11/2020	16/11/2020
							F12	Serious Injury/ Illness/ Impairment	A070504	Premature Discharge of Battery	16/11/2020	16/11/2020
							F19	Surgical Intervention	A070504	Premature Discharge of Battery	16/11/2020	16/11/2020
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	16/11/2020	16/11/2020
					CT3006	Implantable inorganic material	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	16/11/2020	16/11/2020
							F12	Serious Injury/ Illness/ Impairment	A070504	Premature Discharge of Battery	16/11/2020	16/11/2020
							F19	Surgical Intervention	A070504	Premature Discharge of Battery	16/11/2020	16/11/2020
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	16/11/2020	16/11/2020
					CT314	Electrically powered	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	16/11/2020	16/11/2020
							F12	Serious Injury/ Illness/ Impairment	A070504	Premature Discharge of Battery	16/11/2020	16/11/2020
							F19	Surgical Intervention	A070504	Premature Discharge of Battery	16/11/2020	16/11/2020
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	16/11/2020	16/11/2020
					CT321	Long-term surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	16/11/2020	16/11/2020
							F12	Serious Injury/ Illness/ Impairment	A070504	Premature Discharge of Battery	16/11/2020	16/11/2020
							F19	Surgical Intervention	A070504	Premature Discharge of Battery	16/11/2020	16/11/2020
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	16/11/2020	16/11/2020
					CT449	Generators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	16/11/2020	16/11/2020
							F12	Serious Injury/ Illness/ Impairment	A070504	Premature Discharge of Battery	16/11/2020	16/11/2020
							F19	Surgical Intervention	A070504	Premature Discharge of Battery	16/11/2020	16/11/2020
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	16/11/2020	16/11/2020
					CT981	Single use	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	16/11/2020	16/11/2020
							F12	Serious Injury/ Illness/ Impairment	A070504	Premature Discharge of Battery	16/11/2020	16/11/2020
							F19	Surgical Intervention	A070504	Premature Discharge of Battery	16/11/2020	16/11/2020
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	16/11/2020	16/11/2020
					CT983	Surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	16/11/2020	16/11/2020
							F12	Serious Injury/ Illness/ Impairment	A070504	Premature Discharge of Battery	16/11/2020	16/11/2020
							F19	Surgical Intervention	A070504	Premature Discharge of Battery	16/11/2020	16/11/2020
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	16/11/2020	16/11/2020
5973	2020	All other reportable incidents	35852	Single-chamber implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A070909	Oversensing		
							F26	No Health Consequences or Impact	A070909	Oversensing		
5974	2020	All other reportable incidents	47265	Dual-chamber implantable pacemaker, rate-responsive	CT141	Electrodes/Leads and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	24/11/2020	24/11/2020
									A072201	High Impedance	24/11/2020	24/11/2020
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	24/11/2020	24/11/2020
									A072201	High Impedance	24/11/2020	24/11/2020
							F12	Serious Injury/ Illness/ Impairment	A070504	Premature Discharge of Battery	24/11/2020	24/11/2020
									A072201	High Impedance	24/11/2020	24/11/2020
							F15	Recognized Device or Procedural Complication	A070504	Premature Discharge of Battery	24/11/2020	24/11/2020
									A072201	High Impedance	24/11/2020	24/11/2020
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	24/11/2020	24/11/2020
									A072201	High Impedance	24/11/2020	24/11/2020
					CT145	Cardiovascular devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	24/11/2020	24/11/2020
									A072201	High Impedance	24/11/2020	24/11/2020
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	24/11/2020	24/11/2020
									A072201	High Impedance	24/11/2020	24/11/2020
							F12	Serious Injury/ Illness/ Impairment	A070504	Premature Discharge of Battery	24/11/2020	24/11/2020
									A072201	High Impedance	24/11/2020	24/11/2020
							F15	Recognized Device or Procedural Complication	A070504	Premature Discharge of Battery	24/11/2020	24/11/2020
									A072201	High Impedance	24/11/2020	24/11/2020
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	24/11/2020	24/11/2020
									A072201	High Impedance	24/11/2020	24/11/2020
					CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	24/11/2020	24/11/2020
									A072201	High Impedance	24/11/2020	24/11/2020
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	24/11/2020	24/11/2020
									A072201	High Impedance	24/11/2020	24/11/2020
							F12	Serious Injury/ Illness/ Impairment	A070504	Premature Discharge of Battery	24/11/2020	24/11/2020
									A072201	High Impedance	24/11/2020	24/11/2020
							F15	Recognized Device or Procedural Complication	A070504	Premature Discharge of Battery	24/11/2020	24/11/2020
									A072201	High Impedance	24/11/2020	24/11/2020
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	24/11/2020	24/11/2020
									A072201	High Impedance	24/11/2020	24/11/2020
					CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	24/11/2020	24/11/2020
									A072201	High Impedance	24/11/2020	24/11/2020
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	24/11/2020	24/11/2020
									A072201	High Impedance	24/11/2020	24/11/2020
							F12	Serious Injury/ Illness/ Impairment	A070504	Premature Discharge of Battery	24/11/2020	24/11/2020
									A072201	High Impedance	24/11/2020	24/11/2020
							F15	Recognized Device or Procedural Complication	A070504	Premature Discharge of Battery	24/11/2020	24/11/2020
									A072201	High Impedance	24/11/2020	24/11/2020
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	24/11/2020	24/11/2020
									A072201	High Impedance	24/11/2020	24/11/2020
					CT233	Surgical	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	24/11/2020	24/11/2020
									A072201	High Impedance	24/11/2020	24/11/2020
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	24/11/2020	24/11/2020
									A072201	High Impedance	24/11/2020	24/11/2020
							F12	Serious Injury/ Illness/ Impairment	A070504	Premature Discharge of Battery	24/11/2020	24/11/2020

								A072201	High Impedance	24/11/2020	24/11/2020
						F15	Recognized Device or Procedural Complication	A070504	Premature Discharge of Battery	24/11/2020	24/11/2020
								A072201	High Impedance	24/11/2020	24/11/2020
						F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	24/11/2020	24/11/2020
								A072201	High Impedance	24/11/2020	24/11/2020
					CT2406	Implantable	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	24/11/2020	24/11/2020
									A072201	High Impedance	24/11/2020	24/11/2020
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	24/11/2020	24/11/2020
									A072201	High Impedance	24/11/2020	24/11/2020
							F12	Serious Injury/ Illness/ Impairment	A070504	Premature Discharge of Battery	24/11/2020	24/11/2020
									A072201	High Impedance	24/11/2020	24/11/2020
							F15	Recognized Device or Procedural Complication	A070504	Premature Discharge of Battery	24/11/2020	24/11/2020
									A072201	High Impedance	24/11/2020	24/11/2020
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	24/11/2020	24/11/2020
									A072201	High Impedance	24/11/2020	24/11/2020
					CT2407	Active implantable	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	24/11/2020	24/11/2020
									A072201	High Impedance	24/11/2020	24/11/2020
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	24/11/2020	24/11/2020
									A072201	High Impedance	24/11/2020	24/11/2020
							F12	Serious Injury/ Illness/ Impairment	A070504	Premature Discharge of Battery	24/11/2020	24/11/2020
									A072201	High Impedance	24/11/2020	24/11/2020
							F15	Recognized Device or Procedural Complication	A070504	Premature Discharge of Battery	24/11/2020	24/11/2020
									A072201	High Impedance	24/11/2020	24/11/2020
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	24/11/2020	24/11/2020
									A072201	High Impedance	24/11/2020	24/11/2020
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	24/11/2020	24/11/2020
									A072201	High Impedance	24/11/2020	24/11/2020
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	24/11/2020	24/11/2020
									A072201	High Impedance	24/11/2020	24/11/2020
							F12	Serious Injury/ Illness/ Impairment	A070504	Premature Discharge of Battery	24/11/2020	24/11/2020
									A072201	High Impedance	24/11/2020	24/11/2020
							F15	Recognized Device or Procedural Complication	A070504	Premature Discharge of Battery	24/11/2020	24/11/2020
									A072201	High Impedance	24/11/2020	24/11/2020
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	24/11/2020	24/11/2020
									A072201	High Impedance	24/11/2020	24/11/2020
					CT2727	Active implant devices/ancillaries	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	24/11/2020	24/11/2020
									A072201	High Impedance	24/11/2020	24/11/2020
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	24/11/2020	24/11/2020
									A072201	High Impedance	24/11/2020	24/11/2020
							F12	Serious Injury/ Illness/ Impairment	A070504	Premature Discharge of Battery	24/11/2020	24/11/2020
									A072201	High Impedance	24/11/2020	24/11/2020
							F15	Recognized Device or Procedural Complication	A070504	Premature Discharge of Battery	24/11/2020	24/11/2020
									A072201	High Impedance	24/11/2020	24/11/2020
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	24/11/2020	24/11/2020
									A072201	High Impedance	24/11/2020	24/11/2020
					CT2750	Implantable cardiac electrical stimulators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	24/11/2020	24/11/2020
									A072201	High Impedance	24/11/2020	24/11/2020
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	24/11/2020	24/11/2020
									A072201	High Impedance	24/11/2020	24/11/2020
							F12	Serious Injury/ Illness/ Impairment	A070504	Premature Discharge of Battery	24/11/2020	24/11/2020
									A072201	High Impedance	24/11/2020	24/11/2020
							F15	Recognized Device or Procedural Complication	A070504	Premature Discharge of Battery	24/11/2020	24/11/2020
									A072201	High Impedance	24/11/2020	24/11/2020
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	24/11/2020	24/11/2020
									A072201	High Impedance	24/11/2020	24/11/2020
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	24/11/2020	24/11/2020
									A072201	High Impedance	24/11/2020	24/11/2020
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	24/11/2020	24/11/2020
									A072201	High Impedance	24/11/2020	24/11/2020
							F12	Serious Injury/ Illness/ Impairment	A070504	Premature Discharge of Battery	24/11/2020	24/11/2020
									A072201	High Impedance	24/11/2020	24/11/2020
							F15	Recognized Device or Procedural Complication	A070504	Premature Discharge of Battery	24/11/2020	24/11/2020
									A072201	High Impedance	24/11/2020	24/11/2020
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	24/11/2020	24/11/2020
									A072201	High Impedance	24/11/2020	24/11/2020
					CT2985	Active	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	24/11/2020	24/11/2020
									A072201	High Impedance	24/11/2020	24/11/2020
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	24/11/2020	24/11/2020
									A072201	High Impedance	24/11/2020	24/11/2020
							F12	Serious Injury/ Illness/ Impairment	A070504	Premature Discharge of Battery	24/11/2020	24/11/2020
									A072201	High Impedance	24/11/2020	24/11/2020
							F15	Recognized Device or Procedural Complication	A070504	Premature Discharge of Battery	24/11/2020	24/11/2020
									A072201	High Impedance	24/11/2020	24/11/2020
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	24/11/2020	24/11/2020
									A072201	High Impedance	24/11/2020	24/11/2020
					CT3001	Implantable metal	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	24/11/2020	24/11/2020
									A072201	High Impedance	24/11/2020	24/11/2020
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	24/11/2020	24/11/2020
									A072201	High Impedance	24/11/2020	24/11/2020
							F12	Serious Injury/ Illness/ Impairment	A070504	Premature Discharge of Battery	24/11/2020	24/11/2020
									A072201	High Impedance	24/11/2020	24/11/2020
							F15	Recognized Device or Procedural Complication	A070504	Premature Discharge of Battery	24/11/2020	24/11/2020
									A072201	High Impedance	24/11/2020	24/11/2020
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	24/11/2020	24/11/2020
									A072201	High Impedance	24/11/2020	24/11/2020
					CT3002	Implantable synthetic polymer	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	24/11/2020	24/11/2020
									A072201	High Impedance	24/11/2020	24/11/2020
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	24/11/2020	24/11/2020

									A072201	High Impedance	24/11/2020	24/11/2020
							F12	Serious Injury/ Illness/ Impairment	A070504	Premature Discharge of Battery	24/11/2020	24/11/2020
									A072201	High Impedance	24/11/2020	24/11/2020
							F15	Recognized Device or Procedural Complication	A070504	Premature Discharge of Battery	24/11/2020	24/11/2020
									A072201	High Impedance	24/11/2020	24/11/2020
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	24/11/2020	24/11/2020
									A072201	High Impedance	24/11/2020	24/11/2020
					CT3006	Implantable inorganic material	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	24/11/2020	24/11/2020
									A072201	High Impedance	24/11/2020	24/11/2020
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	24/11/2020	24/11/2020
									A072201	High Impedance	24/11/2020	24/11/2020
							F12	Serious Injury/ Illness/ Impairment	A070504	Premature Discharge of Battery	24/11/2020	24/11/2020
									A072201	High Impedance	24/11/2020	24/11/2020
							F15	Recognized Device or Procedural Complication	A070504	Premature Discharge of Battery	24/11/2020	24/11/2020
									A072201	High Impedance	24/11/2020	24/11/2020
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	24/11/2020	24/11/2020
									A072201	High Impedance	24/11/2020	24/11/2020
					CT314	Electrically powered	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	24/11/2020	24/11/2020
									A072201	High Impedance	24/11/2020	24/11/2020
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	24/11/2020	24/11/2020
									A072201	High Impedance	24/11/2020	24/11/2020
							F12	Serious Injury/ Illness/ Impairment	A070504	Premature Discharge of Battery	24/11/2020	24/11/2020
									A072201	High Impedance	24/11/2020	24/11/2020
							F15	Recognized Device or Procedural Complication	A070504	Premature Discharge of Battery	24/11/2020	24/11/2020
									A072201	High Impedance	24/11/2020	24/11/2020
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	24/11/2020	24/11/2020
									A072201	High Impedance	24/11/2020	24/11/2020
					CT321	Long-term surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	24/11/2020	24/11/2020
									A072201	High Impedance	24/11/2020	24/11/2020
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	24/11/2020	24/11/2020
									A072201	High Impedance	24/11/2020	24/11/2020
							F12	Serious Injury/ Illness/ Impairment	A070504	Premature Discharge of Battery	24/11/2020	24/11/2020
									A072201	High Impedance	24/11/2020	24/11/2020
							F15	Recognized Device or Procedural Complication	A070504	Premature Discharge of Battery	24/11/2020	24/11/2020
									A072201	High Impedance	24/11/2020	24/11/2020
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	24/11/2020	24/11/2020
									A072201	High Impedance	24/11/2020	24/11/2020
					CT449	Generators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	24/11/2020	24/11/2020
									A072201	High Impedance	24/11/2020	24/11/2020
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	24/11/2020	24/11/2020
									A072201	High Impedance	24/11/2020	24/11/2020
							F12	Serious Injury/ Illness/ Impairment	A070504	Premature Discharge of Battery	24/11/2020	24/11/2020
									A072201	High Impedance	24/11/2020	24/11/2020
							F15	Recognized Device or Procedural Complication	A070504	Premature Discharge of Battery	24/11/2020	24/11/2020
									A072201	High Impedance	24/11/2020	24/11/2020
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	24/11/2020	24/11/2020
									A072201	High Impedance	24/11/2020	24/11/2020
					CT981	Single use	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	24/11/2020	24/11/2020
									A072201	High Impedance	24/11/2020	24/11/2020
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	24/11/2020	24/11/2020
									A072201	High Impedance	24/11/2020	24/11/2020
							F12	Serious Injury/ Illness/ Impairment	A070504	Premature Discharge of Battery	24/11/2020	24/11/2020
									A072201	High Impedance	24/11/2020	24/11/2020
							F15	Recognized Device or Procedural Complication	A070504	Premature Discharge of Battery	24/11/2020	24/11/2020
									A072201	High Impedance	24/11/2020	24/11/2020
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	24/11/2020	24/11/2020
									A072201	High Impedance	24/11/2020	24/11/2020
					CT983	Surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	24/11/2020	24/11/2020
									A072201	High Impedance	24/11/2020	24/11/2020
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	24/11/2020	24/11/2020
									A072201	High Impedance	24/11/2020	24/11/2020
							F12	Serious Injury/ Illness/ Impairment	A070504	Premature Discharge of Battery	24/11/2020	24/11/2020
									A072201	High Impedance	24/11/2020	24/11/2020
							F15	Recognized Device or Procedural Complication	A070504	Premature Discharge of Battery	24/11/2020	24/11/2020
									A072201	High Impedance	24/11/2020	24/11/2020
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	24/11/2020	24/11/2020
									A072201	High Impedance	24/11/2020	24/11/2020
5975	2022	All other reportable incidents	36241	Pacemaker/defibrillator lead	CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070910	Undersensing		
									A1301	Failure to Read Input Signal		
							F08	Hospitalization or Prolonged Hospitalization	A070910	Undersensing		
									A1301	Failure to Read Input Signal		
							F1905	Device Revision or Replacement	A070910	Undersensing		
									A1301	Failure to Read Input Signal		
5976	2020	All other reportable incidents	37265	Dual-chamber implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A070909	Oversensing		
									A090801	Signal Artifact/Noise		
							F08	Hospitalization or Prolonged Hospitalization	A070909	Oversensing		
									A090801	Signal Artifact/Noise		
							F14	Prolonged Episode of Care	A070909	Oversensing		
									A090801	Signal Artifact/Noise		
5977	2022	All other reportable incidents	47265	Dual-chamber implantable pacemaker, rate-responsive	CT141	Electrodes/Leads and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A0908	Incorrect, Inadequate or Imprecise Result or Readings		
							F26	No Health Consequences or Impact	A0908	Incorrect, Inadequate or Imprecise Result or Readings		
					CT145	Cardiovascular devices	E2403	No Clinical Signs, Symptoms or Conditions	A0908	Incorrect, Inadequate or Imprecise Result or Readings		
							F26	No Health Consequences or Impact	A0908	Incorrect, Inadequate or Imprecise Result or Readings		
					CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A0908	Incorrect, Inadequate or Imprecise Result or Readings		

							F26	No Health Consequences or Impact	A0908	Incorrect, Inadequate or Imprecise Result or Readings		
					CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A0908	Incorrect, Inadequate or Imprecise Result or Readings		
							F26	No Health Consequences or Impact	A0908	Incorrect, Inadequate or Imprecise Result or Readings		
					CT233	Surgical	E2403	No Clinical Signs, Symptoms or Conditions	A0908	Incorrect, Inadequate or Imprecise Result or Readings		
							F26	No Health Consequences or Impact	A0908	Incorrect, Inadequate or Imprecise Result or Readings		
					CT2406	Implantable	E2403	No Clinical Signs, Symptoms or Conditions	A0908	Incorrect, Inadequate or Imprecise Result or Readings		
							F26	No Health Consequences or Impact	A0908	Incorrect, Inadequate or Imprecise Result or Readings		
					CT2407	Active implantable	E2403	No Clinical Signs, Symptoms or Conditions	A0908	Incorrect, Inadequate or Imprecise Result or Readings		
							F26	No Health Consequences or Impact	A0908	Incorrect, Inadequate or Imprecise Result or Readings		
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A0908	Incorrect, Inadequate or Imprecise Result or Readings		
							F26	No Health Consequences or Impact	A0908	Incorrect, Inadequate or Imprecise Result or Readings		
					CT2727	Active implant devices/ancillaries	E2403	No Clinical Signs, Symptoms or Conditions	A0908	Incorrect, Inadequate or Imprecise Result or Readings		
							F26	No Health Consequences or Impact	A0908	Incorrect, Inadequate or Imprecise Result or Readings		
					CT2750	Implantable cardiac electrical stimulators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A0908	Incorrect, Inadequate or Imprecise Result or Readings		
							F26	No Health Consequences or Impact	A0908	Incorrect, Inadequate or Imprecise Result or Readings		
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A0908	Incorrect, Inadequate or Imprecise Result or Readings		
							F26	No Health Consequences or Impact	A0908	Incorrect, Inadequate or Imprecise Result or Readings		
					CT2985	Active	E2403	No Clinical Signs, Symptoms or Conditions	A0908	Incorrect, Inadequate or Imprecise Result or Readings		
							F26	No Health Consequences or Impact	A0908	Incorrect, Inadequate or Imprecise Result or Readings		
					CT3001	Implantable metal	E2403	No Clinical Signs, Symptoms or Conditions	A0908	Incorrect, Inadequate or Imprecise Result or Readings		
							F26	No Health Consequences or Impact	A0908	Incorrect, Inadequate or Imprecise Result or Readings		
					CT3002	Implantable synthetic polymer	E2403	No Clinical Signs, Symptoms or Conditions	A0908	Incorrect, Inadequate or Imprecise Result or Readings		
							F26	No Health Consequences or Impact	A0908	Incorrect, Inadequate or Imprecise Result or Readings		
					CT3006	Implantable inorganic material	E2403	No Clinical Signs, Symptoms or Conditions	A0908	Incorrect, Inadequate or Imprecise Result or Readings		
							F26	No Health Consequences or Impact	A0908	Incorrect, Inadequate or Imprecise Result or Readings		
					CT314	Electrically powered	E2403	No Clinical Signs, Symptoms or Conditions	A0908	Incorrect, Inadequate or Imprecise Result or Readings		
							F26	No Health Consequences or Impact	A0908	Incorrect, Inadequate or Imprecise Result or Readings		
					CT321	Long-term surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A0908	Incorrect, Inadequate or Imprecise Result or Readings		
							F26	No Health Consequences or Impact	A0908	Incorrect, Inadequate or Imprecise Result or Readings		
					CT449	Generators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A0908	Incorrect, Inadequate or Imprecise Result or Readings		
							F26	No Health Consequences or Impact	A0908	Incorrect, Inadequate or Imprecise Result or Readings		
					CT981	Single use	E2403	No Clinical Signs, Symptoms or Conditions	A0908	Incorrect, Inadequate or Imprecise Result or Readings		
							F26	No Health Consequences or Impact	A0908	Incorrect, Inadequate or Imprecise Result or Readings		
					CT983	Surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A0908	Incorrect, Inadequate or Imprecise Result or Readings		
							F26	No Health Consequences or Impact	A0908	Incorrect, Inadequate or Imprecise Result or Readings		
5978	2020	All other reportable incidents	35852	Single-chamber implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A070909	Oversensing		
							F15	Recognized Device or Procedural Complication	A070909	Oversensing		
5979	2020	All other reportable incidents	37265	Dual-chamber implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E2104	Electric Shock	A070909	Oversensing		
									A0712	Pacing Problem		
									A071302	Inappropriate/Inadequate Shock/Stimulation		
							F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A070909	Oversensing		
									A0712	Pacing Problem		
									A071302	Inappropriate/Inadequate Shock/Stimulation		
							F15	Recognized Device or Procedural Complication	A070909	Oversensing		
									A0712	Pacing Problem		
									A071302	Inappropriate/Inadequate Shock/Stimulation		
5980	2020	All other reportable incidents	47265	Dual-chamber implantable pacemaker, rate-responsive	CT141	Electrodes/Leads and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
							F1905	Device Revision or Replacement	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
					CT145	Cardiovascular devices	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
							F1905	Device Revision or Replacement	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
					CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
							F1905	Device Revision or Replacement	A0705	Battery Problem		

									A070504	Premature Discharge of Battery		
						F26	No Health Consequences or Impact	A0705	Battery Problem		
								A070504	Premature Discharge of Battery		
					CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
							F1905	Device Revision or Replacement	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
					CT233	Surgical	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
							F1905	Device Revision or Replacement	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
					CT2406	Implantable	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
							F1905	Device Revision or Replacement	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
					CT2407	Active implantable	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
							F1905	Device Revision or Replacement	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
							F1905	Device Revision or Replacement	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
					CT2727	Active implant devices/ancillaries	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
							F1905	Device Revision or Replacement	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
					CT2750	Implantable cardiac electrical stimulators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
							F1905	Device Revision or Replacement	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
							F1905	Device Revision or Replacement	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
					CT2985	Active	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
							F1905	Device Revision or Replacement	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
					CT3001	Implantable metal	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
							F1905	Device Revision or Replacement	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
					CT3002	Implantable synthetic polymer	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
							F1905	Device Revision or Replacement	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
					CT3006	Implantable inorganic material	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
							F1905	Device Revision or Replacement	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
					CT314	Electrically powered	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
							F1905	Device Revision or Replacement	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
					CT321	Long-term surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
							F1905	Device Revision or Replacement	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
					CT449	Generators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem		

									A070504	Premature Discharge of Battery		
							F1905	Device Revision or Replacement	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
					CT981	Single use	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
							F1905	Device Revision or Replacement	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
					CT983	Surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
							F1905	Device Revision or Replacement	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
5981	2021	All other reportable incidents	47265	Dual-chamber implantable pacemaker, rate-responsive	CT141	Electrodes/Leads and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT145	Cardiovascular devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT233	Surgical	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT2406	Implantable	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT2407	Active implantable	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT2727	Active implant devices/ancillaries	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT2750	Implantable cardiac electrical stimulators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT2985	Active	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT3001	Implantable metal	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT3002	Implantable synthetic polymer	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT3006	Implantable inorganic material	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT314	Electrically powered	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT321	Long-term surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT449	Generators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT981	Single use	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT983	Surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
5982	2021	All other reportable incidents	35852	Single-chamber implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A1107	Data Problem		
							F26	No Health Consequences or Impact	A1107	Data Problem		
5983	2021	All other reportable incidents	47265	Dual-chamber implantable pacemaker, rate-responsive	CT141	Electrodes/Leads and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
							F24	Insufficient Health Impact Information	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
					CT145	Cardiovascular devices	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
							F24	Insufficient Health Impact Information	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
					CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
							F24	Insufficient Health Impact Information	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
					CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
							F24	Insufficient Health Impact Information	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
					CT233	Surgical	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
							F24	Insufficient Health Impact Information	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
					CT2406	Implantable	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
							F24	Insufficient Health Impact Information	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
					CT2407	Active implantable	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
							F24	Insufficient Health Impact Information	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem		

									A070504	Premature Discharge of Battery		
							F24	Insufficient Health Impact Information	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
					CT2727	Active implant devices/ancillaries	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
							F24	Insufficient Health Impact Information	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
					CT2750	Implantable cardiac electrical stimulators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
							F24	Insufficient Health Impact Information	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
							F24	Insufficient Health Impact Information	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
					CT2985	Active	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
							F24	Insufficient Health Impact Information	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
					CT3001	Implantable metal	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
							F24	Insufficient Health Impact Information	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
					CT3002	Implantable synthetic polymer	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
							F24	Insufficient Health Impact Information	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
					CT3006	Implantable inorganic material	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
							F24	Insufficient Health Impact Information	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
					CT314	Electrically powered	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
							F24	Insufficient Health Impact Information	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
					CT321	Long-term surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
							F24	Insufficient Health Impact Information	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
					CT449	Generators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
							F24	Insufficient Health Impact Information	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
					CT981	Single use	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
							F24	Insufficient Health Impact Information	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
					CT983	Surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
							F24	Insufficient Health Impact Information	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
5984	2021	All other reportable incidents	37265	Dual-chamber implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
5985	2021	All other reportable incidents	47267	Single-chamber implantable pacemaker, rate-responsive	CT141	Electrodes/Leads and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A0908	Incorrect, Inadequate or Imprecise Result or Readings	22/06/2021	22/06/2021
							F1905	Device Revision or Replacement	A0908	Incorrect, Inadequate or Imprecise Result or Readings	22/06/2021	22/06/2021
							F24	Insufficient Health Impact Information	A0908	Incorrect, Inadequate or Imprecise Result or Readings	22/06/2021	22/06/2021
					CT145	Cardiovascular devices	E2403	No Clinical Signs, Symptoms or Conditions	A0908	Incorrect, Inadequate or Imprecise Result or Readings	22/06/2021	22/06/2021
							F1905	Device Revision or Replacement	A0908	Incorrect, Inadequate or Imprecise Result or Readings	22/06/2021	22/06/2021
							F24	Insufficient Health Impact Information	A0908	Incorrect, Inadequate or Imprecise Result or Readings	22/06/2021	22/06/2021
					CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A0908	Incorrect, Inadequate or Imprecise Result or Readings	22/06/2021	22/06/2021
							F1905	Device Revision or Replacement	A0908	Incorrect, Inadequate or Imprecise Result or Readings	22/06/2021	22/06/2021
							F24	Insufficient Health Impact Information	A0908	Incorrect, Inadequate or Imprecise Result or Readings	22/06/2021	22/06/2021
					CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A0908	Incorrect, Inadequate or Imprecise Result or Readings	22/06/2021	22/06/2021
							F1905	Device Revision or Replacement	A0908	Incorrect, Inadequate or Imprecise Result or Readings	22/06/2021	22/06/2021
							F24	Insufficient Health Impact Information	A0908	Incorrect, Inadequate or Imprecise Result or Readings	22/06/2021	22/06/2021
					CT233	Surgical	E2403	No Clinical Signs, Symptoms or Conditions	A0908	Incorrect, Inadequate or Imprecise Result or Readings	22/06/2021	22/06/2021
							F1905	Device Revision or Replacement	A0908	Incorrect, Inadequate or Imprecise Result or Readings	22/06/2021	22/06/2021
							F24	Insufficient Health Impact Information	A0908	Incorrect, Inadequate or Imprecise Result or Readings	22/06/2021	22/06/2021
					CT2406	Implantable	E2403	No Clinical Signs, Symptoms or Conditions	A0908	Incorrect, Inadequate or Imprecise Result or Readings	22/06/2021	22/06/2021
							F1905	Device Revision or Replacement	A0908	Incorrect, Inadequate or Imprecise Result or Readings	22/06/2021	22/06/2021
							F24	Insufficient Health Impact Information	A0908	Incorrect, Inadequate or Imprecise Result or Readings	22/06/2021	22/06/2021

					CT2407	Active implantable	E2403	No Clinical Signs, Symptoms or Conditions	A0908	Incorrect, Inadequate or Imprecise Result or Readings	22/06/2021	22/06/2021
							F1905	Device Revision or Replacement	A0908	Incorrect, Inadequate or Imprecise Result or Readings	22/06/2021	22/06/2021
							F24	Insufficient Health Impact Information	A0908	Incorrect, Inadequate or Imprecise Result or Readings	22/06/2021	22/06/2021
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A0908	Incorrect, Inadequate or Imprecise Result or Readings	22/06/2021	22/06/2021
							F1905	Device Revision or Replacement	A0908	Incorrect, Inadequate or Imprecise Result or Readings	22/06/2021	22/06/2021
							F24	Insufficient Health Impact Information	A0908	Incorrect, Inadequate or Imprecise Result or Readings	22/06/2021	22/06/2021
					CT2727	Active implant devices/ancillaries	E2403	No Clinical Signs, Symptoms or Conditions	A0908	Incorrect, Inadequate or Imprecise Result or Readings	22/06/2021	22/06/2021
							F1905	Device Revision or Replacement	A0908	Incorrect, Inadequate or Imprecise Result or Readings	22/06/2021	22/06/2021
							F24	Insufficient Health Impact Information	A0908	Incorrect, Inadequate or Imprecise Result or Readings	22/06/2021	22/06/2021
					CT2750	Implantable cardiac electrical stimulators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A0908	Incorrect, Inadequate or Imprecise Result or Readings	22/06/2021	22/06/2021
							F1905	Device Revision or Replacement	A0908	Incorrect, Inadequate or Imprecise Result or Readings	22/06/2021	22/06/2021
							F24	Insufficient Health Impact Information	A0908	Incorrect, Inadequate or Imprecise Result or Readings	22/06/2021	22/06/2021
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A0908	Incorrect, Inadequate or Imprecise Result or Readings	22/06/2021	22/06/2021
							F1905	Device Revision or Replacement	A0908	Incorrect, Inadequate or Imprecise Result or Readings	22/06/2021	22/06/2021
							F24	Insufficient Health Impact Information	A0908	Incorrect, Inadequate or Imprecise Result or Readings	22/06/2021	22/06/2021
					CT2985	Active	E2403	No Clinical Signs, Symptoms or Conditions	A0908	Incorrect, Inadequate or Imprecise Result or Readings	22/06/2021	22/06/2021
							F1905	Device Revision or Replacement	A0908	Incorrect, Inadequate or Imprecise Result or Readings	22/06/2021	22/06/2021
							F24	Insufficient Health Impact Information	A0908	Incorrect, Inadequate or Imprecise Result or Readings	22/06/2021	22/06/2021
					CT3001	Implantable metal	E2403	No Clinical Signs, Symptoms or Conditions	A0908	Incorrect, Inadequate or Imprecise Result or Readings	22/06/2021	22/06/2021
							F1905	Device Revision or Replacement	A0908	Incorrect, Inadequate or Imprecise Result or Readings	22/06/2021	22/06/2021
							F24	Insufficient Health Impact Information	A0908	Incorrect, Inadequate or Imprecise Result or Readings	22/06/2021	22/06/2021
					CT3002	Implantable synthetic polymer	E2403	No Clinical Signs, Symptoms or Conditions	A0908	Incorrect, Inadequate or Imprecise Result or Readings	22/06/2021	22/06/2021
							F1905	Device Revision or Replacement	A0908	Incorrect, Inadequate or Imprecise Result or Readings	22/06/2021	22/06/2021
							F24	Insufficient Health Impact Information	A0908	Incorrect, Inadequate or Imprecise Result or Readings	22/06/2021	22/06/2021
					CT3006	Implantable inorganic material	E2403	No Clinical Signs, Symptoms or Conditions	A0908	Incorrect, Inadequate or Imprecise Result or Readings	22/06/2021	22/06/2021
							F1905	Device Revision or Replacement	A0908	Incorrect, Inadequate or Imprecise Result or Readings	22/06/2021	22/06/2021
							F24	Insufficient Health Impact Information	A0908	Incorrect, Inadequate or Imprecise Result or Readings	22/06/2021	22/06/2021
					CT314	Electrically powered	E2403	No Clinical Signs, Symptoms or Conditions	A0908	Incorrect, Inadequate or Imprecise Result or Readings	22/06/2021	22/06/2021
							F1905	Device Revision or Replacement	A0908	Incorrect, Inadequate or Imprecise Result or Readings	22/06/2021	22/06/2021
							F24	Insufficient Health Impact Information	A0908	Incorrect, Inadequate or Imprecise Result or Readings	22/06/2021	22/06/2021
					CT321	Long-term surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A0908	Incorrect, Inadequate or Imprecise Result or Readings	22/06/2021	22/06/2021
							F1905	Device Revision or Replacement	A0908	Incorrect, Inadequate or Imprecise Result or Readings	22/06/2021	22/06/2021
							F24	Insufficient Health Impact Information	A0908	Incorrect, Inadequate or Imprecise Result or Readings	22/06/2021	22/06/2021
					CT449	Generators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A0908	Incorrect, Inadequate or Imprecise Result or Readings	22/06/2021	22/06/2021
							F1905	Device Revision or Replacement	A0908	Incorrect, Inadequate or Imprecise Result or Readings	22/06/2021	22/06/2021
							F24	Insufficient Health Impact Information	A0908	Incorrect, Inadequate or Imprecise Result or Readings	22/06/2021	22/06/2021
					CT981	Single use	E2403	No Clinical Signs, Symptoms or Conditions	A0908	Incorrect, Inadequate or Imprecise Result or Readings	22/06/2021	22/06/2021
							F1905	Device Revision or Replacement	A0908	Incorrect, Inadequate or Imprecise Result or Readings	22/06/2021	22/06/2021
							F24	Insufficient Health Impact Information	A0908	Incorrect, Inadequate or Imprecise Result or Readings	22/06/2021	22/06/2021
					CT983	Surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A0908	Incorrect, Inadequate or Imprecise Result or Readings	22/06/2021	22/06/2021
							F1905	Device Revision or Replacement	A0908	Incorrect, Inadequate or Imprecise Result or Readings	22/06/2021	22/06/2021
							F24	Insufficient Health Impact Information	A0908	Incorrect, Inadequate or Imprecise Result or Readings	22/06/2021	22/06/2021
5986	2021	All other reportable incidents	35852	Single-chamber implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem	05/01/2022	05/01/2022
									A070504	Premature Discharge of Battery	05/01/2022	05/01/2022
							F08	Hospitalization or Prolonged Hospitalization	A0705	Battery Problem	05/01/2022	05/01/2022
									A070504	Premature Discharge of Battery	05/01/2022	05/01/2022
							F1903	Device Explantation	A0705	Battery Problem	05/01/2022	05/01/2022
									A070504	Premature Discharge of Battery	05/01/2022	05/01/2022
5987	2021	All other reportable incidents	47265	Dual-chamber implantable pacemaker, rate-responsive	CT141	Electrodes/Leads and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem	05/07/2021	05/07/2021
									A071102	Failure to Interrogate	05/07/2021	05/07/2021
									A1605	Reset Problem	05/07/2021	05/07/2021
							F1905	Device Revision or Replacement	A0705	Battery Problem	05/07/2021	05/07/2021
									A071102	Failure to Interrogate	05/07/2021	05/07/2021
									A1605	Reset Problem	05/07/2021	05/07/2021
							F26	No Health Consequences or Impact	A0705	Battery Problem	05/07/2021	05/07/2021
									A071102	Failure to Interrogate	05/07/2021	05/07/2021

								A1605	Reset Problem	05/07/2021	05/07/2021
					CT145	Cardiovascular devices	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem	05/07/2021	05/07/2021
									A071102	Failure to Interrogate	05/07/2021	05/07/2021
									A1605	Reset Problem	05/07/2021	05/07/2021
							F1905	Device Revision or Replacement	A0705	Battery Problem	05/07/2021	05/07/2021
									A071102	Failure to Interrogate	05/07/2021	05/07/2021
									A1605	Reset Problem	05/07/2021	05/07/2021
							F26	No Health Consequences or Impact	A0705	Battery Problem	05/07/2021	05/07/2021
									A071102	Failure to Interrogate	05/07/2021	05/07/2021
									A1605	Reset Problem	05/07/2021	05/07/2021
					CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem	05/07/2021	05/07/2021
									A071102	Failure to Interrogate	05/07/2021	05/07/2021
									A1605	Reset Problem	05/07/2021	05/07/2021
							F1905	Device Revision or Replacement	A0705	Battery Problem	05/07/2021	05/07/2021
									A071102	Failure to Interrogate	05/07/2021	05/07/2021
									A1605	Reset Problem	05/07/2021	05/07/2021
							F26	No Health Consequences or Impact	A0705	Battery Problem	05/07/2021	05/07/2021
									A071102	Failure to Interrogate	05/07/2021	05/07/2021
									A1605	Reset Problem	05/07/2021	05/07/2021
					CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem	05/07/2021	05/07/2021
									A071102	Failure to Interrogate	05/07/2021	05/07/2021
									A1605	Reset Problem	05/07/2021	05/07/2021
							F1905	Device Revision or Replacement	A0705	Battery Problem	05/07/2021	05/07/2021
									A071102	Failure to Interrogate	05/07/2021	05/07/2021
									A1605	Reset Problem	05/07/2021	05/07/2021
							F26	No Health Consequences or Impact	A0705	Battery Problem	05/07/2021	05/07/2021
									A071102	Failure to Interrogate	05/07/2021	05/07/2021
									A1605	Reset Problem	05/07/2021	05/07/2021
					CT233	Surgical	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem	05/07/2021	05/07/2021
									A071102	Failure to Interrogate	05/07/2021	05/07/2021
									A1605	Reset Problem	05/07/2021	05/07/2021
							F1905	Device Revision or Replacement	A0705	Battery Problem	05/07/2021	05/07/2021
									A071102	Failure to Interrogate	05/07/2021	05/07/2021
									A1605	Reset Problem	05/07/2021	05/07/2021
							F26	No Health Consequences or Impact	A0705	Battery Problem	05/07/2021	05/07/2021
									A071102	Failure to Interrogate	05/07/2021	05/07/2021
									A1605	Reset Problem	05/07/2021	05/07/2021
					CT2406	Implantable	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem	05/07/2021	05/07/2021
									A071102	Failure to Interrogate	05/07/2021	05/07/2021
									A1605	Reset Problem	05/07/2021	05/07/2021
							F1905	Device Revision or Replacement	A0705	Battery Problem	05/07/2021	05/07/2021
									A071102	Failure to Interrogate	05/07/2021	05/07/2021
									A1605	Reset Problem	05/07/2021	05/07/2021
							F26	No Health Consequences or Impact	A0705	Battery Problem	05/07/2021	05/07/2021
									A071102	Failure to Interrogate	05/07/2021	05/07/2021
									A1605	Reset Problem	05/07/2021	05/07/2021
					CT2407	Active implantable	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem	05/07/2021	05/07/2021
									A071102	Failure to Interrogate	05/07/2021	05/07/2021
									A1605	Reset Problem	05/07/2021	05/07/2021
							F1905	Device Revision or Replacement	A0705	Battery Problem	05/07/2021	05/07/2021
									A071102	Failure to Interrogate	05/07/2021	05/07/2021
									A1605	Reset Problem	05/07/2021	05/07/2021
							F26	No Health Consequences or Impact	A0705	Battery Problem	05/07/2021	05/07/2021
									A071102	Failure to Interrogate	05/07/2021	05/07/2021
									A1605	Reset Problem	05/07/2021	05/07/2021
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem	05/07/2021	05/07/2021
									A071102	Failure to Interrogate	05/07/2021	05/07/2021
									A1605	Reset Problem	05/07/2021	05/07/2021
							F1905	Device Revision or Replacement	A0705	Battery Problem	05/07/2021	05/07/2021
									A071102	Failure to Interrogate	05/07/2021	05/07/2021
									A1605	Reset Problem	05/07/2021	05/07/2021
							F26	No Health Consequences or Impact	A0705	Battery Problem	05/07/2021	05/07/2021
									A071102	Failure to Interrogate	05/07/2021	05/07/2021
									A1605	Reset Problem	05/07/2021	05/07/2021
					CT2727	Active implant devices/ancillaries	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem	05/07/2021	05/07/2021
									A071102	Failure to Interrogate	05/07/2021	05/07/2021
									A1605	Reset Problem	05/07/2021	05/07/2021
							F1905	Device Revision or Replacement	A0705	Battery Problem	05/07/2021	05/07/2021
									A071102	Failure to Interrogate	05/07/2021	05/07/2021
									A1605	Reset Problem	05/07/2021	05/07/2021
							F26	No Health Consequences or Impact	A0705	Battery Problem	05/07/2021	05/07/2021
									A071102	Failure to Interrogate	05/07/2021	05/07/2021
									A1605	Reset Problem	05/07/2021	05/07/2021
					CT2750	Implantable cardiac electrical stimulators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem	05/07/2021	05/07/2021
									A071102	Failure to Interrogate	05/07/2021	05/07/2021
									A1605	Reset Problem	05/07/2021	05/07/2021
							F1905	Device Revision or Replacement	A0705	Battery Problem	05/07/2021	05/07/2021
									A071102	Failure to Interrogate	05/07/2021	05/07/2021
									A1605	Reset Problem	05/07/2021	05/07/2021
							F26	No Health Consequences or Impact	A0705	Battery Problem	05/07/2021	05/07/2021
									A071102	Failure to Interrogate	05/07/2021	05/07/2021
									A1605	Reset Problem	05/07/2021	05/07/2021
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem	05/07/2021	05/07/2021
									A071102	Failure to Interrogate	05/07/2021	05/07/2021
									A1605	Reset Problem	05/07/2021	05/07/2021
							F1905	Device Revision or Replacement	A0705	Battery Problem	05/07/2021	05/07/2021
									A071102	Failure to Interrogate	05/07/2021	05/07/2021
									A1605	Reset Problem	05/07/2021	05/07/2021

							F26	No Health Consequences or Impact	A0705	Battery Problem	05/07/2021	05/07/2021
									A071102	Failure to Interrogate	05/07/2021	05/07/2021
									A1605	Reset Problem	05/07/2021	05/07/2021
					CT2985	Active	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem	05/07/2021	05/07/2021
									A071102	Failure to Interrogate	05/07/2021	05/07/2021
									A1605	Reset Problem	05/07/2021	05/07/2021
							F1905	Device Revision or Replacement	A0705	Battery Problem	05/07/2021	05/07/2021
									A071102	Failure to Interrogate	05/07/2021	05/07/2021
									A1605	Reset Problem	05/07/2021	05/07/2021
							F26	No Health Consequences or Impact	A0705	Battery Problem	05/07/2021	05/07/2021
									A071102	Failure to Interrogate	05/07/2021	05/07/2021
									A1605	Reset Problem	05/07/2021	05/07/2021
					CT3001	Implantable metal	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem	05/07/2021	05/07/2021
									A071102	Failure to Interrogate	05/07/2021	05/07/2021
									A1605	Reset Problem	05/07/2021	05/07/2021
							F1905	Device Revision or Replacement	A0705	Battery Problem	05/07/2021	05/07/2021
									A071102	Failure to Interrogate	05/07/2021	05/07/2021
									A1605	Reset Problem	05/07/2021	05/07/2021
							F26	No Health Consequences or Impact	A0705	Battery Problem	05/07/2021	05/07/2021
									A071102	Failure to Interrogate	05/07/2021	05/07/2021
									A1605	Reset Problem	05/07/2021	05/07/2021
					CT3002	Implantable synthetic polymer	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem	05/07/2021	05/07/2021
									A071102	Failure to Interrogate	05/07/2021	05/07/2021
									A1605	Reset Problem	05/07/2021	05/07/2021
							F1905	Device Revision or Replacement	A0705	Battery Problem	05/07/2021	05/07/2021
									A071102	Failure to Interrogate	05/07/2021	05/07/2021
									A1605	Reset Problem	05/07/2021	05/07/2021
							F26	No Health Consequences or Impact	A0705	Battery Problem	05/07/2021	05/07/2021
									A071102	Failure to Interrogate	05/07/2021	05/07/2021
									A1605	Reset Problem	05/07/2021	05/07/2021
					CT3006	Implantable inorganic material	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem	05/07/2021	05/07/2021
									A071102	Failure to Interrogate	05/07/2021	05/07/2021
									A1605	Reset Problem	05/07/2021	05/07/2021
							F1905	Device Revision or Replacement	A0705	Battery Problem	05/07/2021	05/07/2021
									A071102	Failure to Interrogate	05/07/2021	05/07/2021
									A1605	Reset Problem	05/07/2021	05/07/2021
							F26	No Health Consequences or Impact	A0705	Battery Problem	05/07/2021	05/07/2021
									A071102	Failure to Interrogate	05/07/2021	05/07/2021
									A1605	Reset Problem	05/07/2021	05/07/2021
					CT314	Electrically powered	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem	05/07/2021	05/07/2021
									A071102	Failure to Interrogate	05/07/2021	05/07/2021
									A1605	Reset Problem	05/07/2021	05/07/2021
							F1905	Device Revision or Replacement	A0705	Battery Problem	05/07/2021	05/07/2021
									A071102	Failure to Interrogate	05/07/2021	05/07/2021
									A1605	Reset Problem	05/07/2021	05/07/2021
							F26	No Health Consequences or Impact	A0705	Battery Problem	05/07/2021	05/07/2021
									A071102	Failure to Interrogate	05/07/2021	05/07/2021
									A1605	Reset Problem	05/07/2021	05/07/2021
					CT321	Long-term surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem	05/07/2021	05/07/2021
									A071102	Failure to Interrogate	05/07/2021	05/07/2021
									A1605	Reset Problem	05/07/2021	05/07/2021
							F1905	Device Revision or Replacement	A0705	Battery Problem	05/07/2021	05/07/2021
									A071102	Failure to Interrogate	05/07/2021	05/07/2021
									A1605	Reset Problem	05/07/2021	05/07/2021
							F26	No Health Consequences or Impact	A0705	Battery Problem	05/07/2021	05/07/2021
									A071102	Failure to Interrogate	05/07/2021	05/07/2021
									A1605	Reset Problem	05/07/2021	05/07/2021
					CT449	Generators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem	05/07/2021	05/07/2021
									A071102	Failure to Interrogate	05/07/2021	05/07/2021
									A1605	Reset Problem	05/07/2021	05/07/2021
							F1905	Device Revision or Replacement	A0705	Battery Problem	05/07/2021	05/07/2021
									A071102	Failure to Interrogate	05/07/2021	05/07/2021
									A1605	Reset Problem	05/07/2021	05/07/2021
							F26	No Health Consequences or Impact	A0705	Battery Problem	05/07/2021	05/07/2021
									A071102	Failure to Interrogate	05/07/2021	05/07/2021
									A1605	Reset Problem	05/07/2021	05/07/2021
					CT981	Single use	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem	05/07/2021	05/07/2021
									A071102	Failure to Interrogate	05/07/2021	05/07/2021
									A1605	Reset Problem	05/07/2021	05/07/2021
							F1905	Device Revision or Replacement	A0705	Battery Problem	05/07/2021	05/07/2021
									A071102	Failure to Interrogate	05/07/2021	05/07/2021
									A1605	Reset Problem	05/07/2021	05/07/2021
							F26	No Health Consequences or Impact	A0705	Battery Problem	05/07/2021	05/07/2021
									A071102	Failure to Interrogate	05/07/2021	05/07/2021
									A1605	Reset Problem	05/07/2021	05/07/2021
					CT983	Surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem	05/07/2021	05/07/2021
									A071102	Failure to Interrogate	05/07/2021	05/07/2021
									A1605	Reset Problem	05/07/2021	05/07/2021
							F1905	Device Revision or Replacement	A0705	Battery Problem	05/07/2021	05/07/2021
									A071102	Failure to Interrogate	05/07/2021	05/07/2021
									A1605	Reset Problem	05/07/2021	05/07/2021
							F26	No Health Consequences or Impact	A0705	Battery Problem	05/07/2021	05/07/2021
									A071102	Failure to Interrogate	05/07/2021	05/07/2021
									A1605	Reset Problem	05/07/2021	05/07/2021
5988	2020	All other reportable incidents	17882	Physiologic-monitoring defibrillation system	CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070803	Failure to Power Up		
							F27	Problem Identified During Non-Clinical Procedure	A070803	Failure to Power Up		
5989	2021	All other reportable incidents	35852	Single-chamber implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	28/07/2021	28/07/2021
							F19	Surgical Intervention	A070504	Premature Discharge of Battery	28/07/2021	28/07/2021

5990	2021	All other reportable incidents	35852	Single-chamber implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	20/09/2021	20/09/2021
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery	20/09/2021	20/09/2021
5991	2022	All other reportable incidents	47263	Cardiac resynchronization therapy implantable pacemaker	CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A0908	Incorrect, Inadequate or Imprecise Result or Readings	09/02/2022	09/02/2022
							F08	Hospitalization or Prolonged Hospitalization	A0908	Incorrect, Inadequate or Imprecise Result or Readings	09/02/2022	09/02/2022
							F1905	Device Revision or Replacement	A0908	Incorrect, Inadequate or Imprecise Result or Readings	09/02/2022	09/02/2022
					CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A0908	Incorrect, Inadequate or Imprecise Result or Readings	09/02/2022	09/02/2022
							F08	Hospitalization or Prolonged Hospitalization	A0908	Incorrect, Inadequate or Imprecise Result or Readings	09/02/2022	09/02/2022
							F1905	Device Revision or Replacement	A0908	Incorrect, Inadequate or Imprecise Result or Readings	09/02/2022	09/02/2022
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A0908	Incorrect, Inadequate or Imprecise Result or Readings	09/02/2022	09/02/2022
							F08	Hospitalization or Prolonged Hospitalization	A0908	Incorrect, Inadequate or Imprecise Result or Readings	09/02/2022	09/02/2022
							F1905	Device Revision or Replacement	A0908	Incorrect, Inadequate or Imprecise Result or Readings	09/02/2022	09/02/2022
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A0908	Incorrect, Inadequate or Imprecise Result or Readings	09/02/2022	09/02/2022
							F08	Hospitalization or Prolonged Hospitalization	A0908	Incorrect, Inadequate or Imprecise Result or Readings	09/02/2022	09/02/2022
							F1905	Device Revision or Replacement	A0908	Incorrect, Inadequate or Imprecise Result or Readings	09/02/2022	09/02/2022
5992	2022	All other reportable incidents	35852	Single-chamber implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E2104	Electric Shock	A071302	Inappropriate/Inadequate Shock/Stimulation	02/02/2022	02/02/2022
									A0908	Incorrect, Inadequate or Imprecise Result or Readings	02/02/2022	02/02/2022
							E2330	Pain	A071302	Inappropriate/Inadequate Shock/Stimulation	02/02/2022	02/02/2022
									A0908	Incorrect, Inadequate or Imprecise Result or Readings	02/02/2022	02/02/2022
							E2403	No Clinical Signs, Symptoms or Conditions	A071302	Inappropriate/Inadequate Shock/Stimulation	02/02/2022	02/02/2022
									A0908	Incorrect, Inadequate or Imprecise Result or Readings	02/02/2022	02/02/2022
							F08	Hospitalization or Prolonged Hospitalization	A071302	Inappropriate/Inadequate Shock/Stimulation	02/02/2022	02/02/2022
									A0908	Incorrect, Inadequate or Imprecise Result or Readings	02/02/2022	02/02/2022
							F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A071302	Inappropriate/Inadequate Shock/Stimulation	02/02/2022	02/02/2022
									A0908	Incorrect, Inadequate or Imprecise Result or Readings	02/02/2022	02/02/2022
							F1905	Device Revision or Replacement	A071302	Inappropriate/Inadequate Shock/Stimulation	02/02/2022	02/02/2022
									A0908	Incorrect, Inadequate or Imprecise Result or Readings	02/02/2022	02/02/2022
5993	2020	All other reportable incidents	36241	Pacemaker/defibrillator lead	CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070101	Failure to Capture	30/09/2020	30/09/2020
									A070102	High Capture Threshold	30/09/2020	30/09/2020
							F19	Surgical Intervention	A070101	Failure to Capture	30/09/2020	30/09/2020
									A070102	High Capture Threshold	30/09/2020	30/09/2020
							F1905	Device Revision or Replacement	A070101	Failure to Capture	30/09/2020	30/09/2020
									A070102	High Capture Threshold	30/09/2020	30/09/2020
5994	2020	All other reportable incidents	36241	Pacemaker/defibrillator lead	CT1684	Cardiac pacemakers and associated devices	E1906	Unspecified Infection	A24	No Apparent Device Problem	12/10/2020	12/10/2020
							F1905	Device Revision or Replacement	A24	No Apparent Device Problem	12/10/2020	12/10/2020
5995	2020	All other reportable incidents	36241	Pacemaker/defibrillator lead	CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070101	Failure to Capture		
									A070909	Oversensing		
									A0712	Pacing Problem		
							F26	No Health Consequences or Impact	A070101	Failure to Capture		
									A070909	Oversensing		
									A0712	Pacing Problem		
5996	2022	All other reportable incidents	47267	Single-chamber implantable pacemaker, rate-responsive	CT141	Electrodes/Leads and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A0908	Incorrect, Inadequate or Imprecise Result or Readings	17/02/2022	17/02/2022
							F08	Hospitalization or Prolonged Hospitalization	A0908	Incorrect, Inadequate or Imprecise Result or Readings	17/02/2022	17/02/2022
							F1905	Device Revision or Replacement	A0908	Incorrect, Inadequate or Imprecise Result or Readings	17/02/2022	17/02/2022
					CT145	Cardiovascular devices	E2403	No Clinical Signs, Symptoms or Conditions	A0908	Incorrect, Inadequate or Imprecise Result or Readings	17/02/2022	17/02/2022
							F08	Hospitalization or Prolonged Hospitalization	A0908	Incorrect, Inadequate or Imprecise Result or Readings	17/02/2022	17/02/2022
							F1905	Device Revision or Replacement	A0908	Incorrect, Inadequate or Imprecise Result or Readings	17/02/2022	17/02/2022
					CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A0908	Incorrect, Inadequate or Imprecise Result or Readings	17/02/2022	17/02/2022
							F08	Hospitalization or Prolonged Hospitalization	A0908	Incorrect, Inadequate or Imprecise Result or Readings	17/02/2022	17/02/2022
							F1905	Device Revision or Replacement	A0908	Incorrect, Inadequate or Imprecise Result or Readings	17/02/2022	17/02/2022
					CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A0908	Incorrect, Inadequate or Imprecise Result or Readings	17/02/2022	17/02/2022
							F08	Hospitalization or Prolonged Hospitalization	A0908	Incorrect, Inadequate or Imprecise Result or Readings	17/02/2022	17/02/2022
							F1905	Device Revision or Replacement	A0908	Incorrect, Inadequate or Imprecise Result or Readings	17/02/2022	17/02/2022
					CT233	Surgical	E2403	No Clinical Signs, Symptoms or Conditions	A0908	Incorrect, Inadequate or Imprecise Result or Readings	17/02/2022	17/02/2022
							F08	Hospitalization or Prolonged Hospitalization	A0908	Incorrect, Inadequate or Imprecise Result or Readings	17/02/2022	17/02/2022
							F1905	Device Revision or Replacement	A0908	Incorrect, Inadequate or Imprecise Result or Readings	17/02/2022	17/02/2022
					CT2406	Implantable	E2403	No Clinical Signs, Symptoms or Conditions	A0908	Incorrect, Inadequate or Imprecise Result or Readings	17/02/2022	17/02/2022
							F08	Hospitalization or Prolonged Hospitalization	A0908	Incorrect, Inadequate or Imprecise Result or Readings	17/02/2022	17/02/2022
							F1905	Device Revision or Replacement	A0908	Incorrect, Inadequate or Imprecise Result or Readings	17/02/2022	17/02/2022
					CT2407	Active implantable	E2403	No Clinical Signs, Symptoms or Conditions	A0908	Incorrect, Inadequate or Imprecise Result or Readings	17/02/2022	17/02/2022

							F08	Hospitalization or Prolonged Hospitalization	A0908	Incorrect, Inadequate or Imprecise Result or Readings	17/02/2022	17/02/2022
							F1905	Device Revision or Replacement	A0908	Incorrect, Inadequate or Imprecise Result or Readings	17/02/2022	17/02/2022
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A0908	Incorrect, Inadequate or Imprecise Result or Readings	17/02/2022	17/02/2022
							F08	Hospitalization or Prolonged Hospitalization	A0908	Incorrect, Inadequate or Imprecise Result or Readings	17/02/2022	17/02/2022
							F1905	Device Revision or Replacement	A0908	Incorrect, Inadequate or Imprecise Result or Readings	17/02/2022	17/02/2022
					CT2727	Active implant devices/ancillaries	E2403	No Clinical Signs, Symptoms or Conditions	A0908	Incorrect, Inadequate or Imprecise Result or Readings	17/02/2022	17/02/2022
							F08	Hospitalization or Prolonged Hospitalization	A0908	Incorrect, Inadequate or Imprecise Result or Readings	17/02/2022	17/02/2022
							F1905	Device Revision or Replacement	A0908	Incorrect, Inadequate or Imprecise Result or Readings	17/02/2022	17/02/2022
					CT2750	Implantable cardiac electrical stimulators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A0908	Incorrect, Inadequate or Imprecise Result or Readings	17/02/2022	17/02/2022
							F08	Hospitalization or Prolonged Hospitalization	A0908	Incorrect, Inadequate or Imprecise Result or Readings	17/02/2022	17/02/2022
							F1905	Device Revision or Replacement	A0908	Incorrect, Inadequate or Imprecise Result or Readings	17/02/2022	17/02/2022
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A0908	Incorrect, Inadequate or Imprecise Result or Readings	17/02/2022	17/02/2022
							F08	Hospitalization or Prolonged Hospitalization	A0908	Incorrect, Inadequate or Imprecise Result or Readings	17/02/2022	17/02/2022
							F1905	Device Revision or Replacement	A0908	Incorrect, Inadequate or Imprecise Result or Readings	17/02/2022	17/02/2022
					CT2985	Active	E2403	No Clinical Signs, Symptoms or Conditions	A0908	Incorrect, Inadequate or Imprecise Result or Readings	17/02/2022	17/02/2022
							F08	Hospitalization or Prolonged Hospitalization	A0908	Incorrect, Inadequate or Imprecise Result or Readings	17/02/2022	17/02/2022
							F1905	Device Revision or Replacement	A0908	Incorrect, Inadequate or Imprecise Result or Readings	17/02/2022	17/02/2022
					CT3001	Implantable metal	E2403	No Clinical Signs, Symptoms or Conditions	A0908	Incorrect, Inadequate or Imprecise Result or Readings	17/02/2022	17/02/2022
							F08	Hospitalization or Prolonged Hospitalization	A0908	Incorrect, Inadequate or Imprecise Result or Readings	17/02/2022	17/02/2022
							F1905	Device Revision or Replacement	A0908	Incorrect, Inadequate or Imprecise Result or Readings	17/02/2022	17/02/2022
					CT3002	Implantable synthetic polymer	E2403	No Clinical Signs, Symptoms or Conditions	A0908	Incorrect, Inadequate or Imprecise Result or Readings	17/02/2022	17/02/2022
							F08	Hospitalization or Prolonged Hospitalization	A0908	Incorrect, Inadequate or Imprecise Result or Readings	17/02/2022	17/02/2022
							F1905	Device Revision or Replacement	A0908	Incorrect, Inadequate or Imprecise Result or Readings	17/02/2022	17/02/2022
					CT3006	Implantable inorganic material	E2403	No Clinical Signs, Symptoms or Conditions	A0908	Incorrect, Inadequate or Imprecise Result or Readings	17/02/2022	17/02/2022
							F08	Hospitalization or Prolonged Hospitalization	A0908	Incorrect, Inadequate or Imprecise Result or Readings	17/02/2022	17/02/2022
							F1905	Device Revision or Replacement	A0908	Incorrect, Inadequate or Imprecise Result or Readings	17/02/2022	17/02/2022
					CT314	Electrically powered	E2403	No Clinical Signs, Symptoms or Conditions	A0908	Incorrect, Inadequate or Imprecise Result or Readings	17/02/2022	17/02/2022
							F08	Hospitalization or Prolonged Hospitalization	A0908	Incorrect, Inadequate or Imprecise Result or Readings	17/02/2022	17/02/2022
							F1905	Device Revision or Replacement	A0908	Incorrect, Inadequate or Imprecise Result or Readings	17/02/2022	17/02/2022
					CT321	Long-term surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A0908	Incorrect, Inadequate or Imprecise Result or Readings	17/02/2022	17/02/2022
							F08	Hospitalization or Prolonged Hospitalization	A0908	Incorrect, Inadequate or Imprecise Result or Readings	17/02/2022	17/02/2022
							F1905	Device Revision or Replacement	A0908	Incorrect, Inadequate or Imprecise Result or Readings	17/02/2022	17/02/2022
					CT449	Generators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A0908	Incorrect, Inadequate or Imprecise Result or Readings	17/02/2022	17/02/2022
							F08	Hospitalization or Prolonged Hospitalization	A0908	Incorrect, Inadequate or Imprecise Result or Readings	17/02/2022	17/02/2022
							F1905	Device Revision or Replacement	A0908	Incorrect, Inadequate or Imprecise Result or Readings	17/02/2022	17/02/2022
					CT981	Single use	E2403	No Clinical Signs, Symptoms or Conditions	A0908	Incorrect, Inadequate or Imprecise Result or Readings	17/02/2022	17/02/2022
							F08	Hospitalization or Prolonged Hospitalization	A0908	Incorrect, Inadequate or Imprecise Result or Readings	17/02/2022	17/02/2022
							F1905	Device Revision or Replacement	A0908	Incorrect, Inadequate or Imprecise Result or Readings	17/02/2022	17/02/2022
					CT983	Surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A0908	Incorrect, Inadequate or Imprecise Result or Readings	17/02/2022	17/02/2022
							F08	Hospitalization or Prolonged Hospitalization	A0908	Incorrect, Inadequate or Imprecise Result or Readings	17/02/2022	17/02/2022
							F1905	Device Revision or Replacement	A0908	Incorrect, Inadequate or Imprecise Result or Readings	17/02/2022	17/02/2022
5997	2022	All other reportable incidents	37265	Dual-chamber implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E2104	Electric Shock	A071302	Inappropriate/Inadequate Shock/Stimulation		
							E2403	No Clinical Signs, Symptoms or Conditions	A071302	Inappropriate/Inadequate Shock/Stimulation		
							F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A071302	Inappropriate/Inadequate Shock/Stimulation		
							F15	Recognized Device or Procedural Complication	A071302	Inappropriate/Inadequate Shock/Stimulation		
5998	2022	All other reportable incidents	35852	Single-chamber implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem	16/05/2022	16/05/2022
									A070504	Premature Discharge of Battery	16/05/2022	16/05/2022
							F08	Hospitalization or Prolonged Hospitalization	A0705	Battery Problem	16/05/2022	16/05/2022
									A070504	Premature Discharge of Battery	16/05/2022	16/05/2022
							F1905	Device Revision or Replacement	A0705	Battery Problem	16/05/2022	16/05/2022
									A070504	Premature Discharge of Battery	16/05/2022	16/05/2022
5999	2021	All other reportable incidents	60789	Intracardiac pacemaker	CT130	Catheters and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A23	Use of Device Problem	26/05/2021	26/05/2021
							F1905	Device Revision or Replacement	A23	Use of Device Problem	26/05/2021	26/05/2021
							F26	No Health Consequences or Impact	A23	Use of Device Problem	26/05/2021	26/05/2021
					CT145	Cardiovascular devices	E2403	No Clinical Signs, Symptoms or Conditions	A23	Use of Device Problem	26/05/2021	26/05/2021
							F1905	Device Revision or Replacement	A23	Use of Device Problem	26/05/2021	26/05/2021

							F26	No Health Consequences or Impact	A23	Use of Device Problem	26/05/2021	26/05/2021
					CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A23	Use of Device Problem	26/05/2021	26/05/2021
							F1905	Device Revision or Replacement	A23	Use of Device Problem	26/05/2021	26/05/2021
							F26	No Health Consequences or Impact	A23	Use of Device Problem	26/05/2021	26/05/2021
					CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A23	Use of Device Problem	26/05/2021	26/05/2021
							F1905	Device Revision or Replacement	A23	Use of Device Problem	26/05/2021	26/05/2021
							F26	No Health Consequences or Impact	A23	Use of Device Problem	26/05/2021	26/05/2021
					CT233	Surgical	E2403	No Clinical Signs, Symptoms or Conditions	A23	Use of Device Problem	26/05/2021	26/05/2021
							F1905	Device Revision or Replacement	A23	Use of Device Problem	26/05/2021	26/05/2021
							F26	No Health Consequences or Impact	A23	Use of Device Problem	26/05/2021	26/05/2021
					CT2406	Implantable	E2403	No Clinical Signs, Symptoms or Conditions	A23	Use of Device Problem	26/05/2021	26/05/2021
							F1905	Device Revision or Replacement	A23	Use of Device Problem	26/05/2021	26/05/2021
							F26	No Health Consequences or Impact	A23	Use of Device Problem	26/05/2021	26/05/2021
					CT2407	Active implantable	E2403	No Clinical Signs, Symptoms or Conditions	A23	Use of Device Problem	26/05/2021	26/05/2021
							F1905	Device Revision or Replacement	A23	Use of Device Problem	26/05/2021	26/05/2021
							F26	No Health Consequences or Impact	A23	Use of Device Problem	26/05/2021	26/05/2021
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A23	Use of Device Problem	26/05/2021	26/05/2021
							F1905	Device Revision or Replacement	A23	Use of Device Problem	26/05/2021	26/05/2021
							F26	No Health Consequences or Impact	A23	Use of Device Problem	26/05/2021	26/05/2021
					CT2727	Active implant devices/ancillaries	E2403	No Clinical Signs, Symptoms or Conditions	A23	Use of Device Problem	26/05/2021	26/05/2021
							F1905	Device Revision or Replacement	A23	Use of Device Problem	26/05/2021	26/05/2021
							F26	No Health Consequences or Impact	A23	Use of Device Problem	26/05/2021	26/05/2021
					CT2750	Implantable cardiac electrical stimulators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A23	Use of Device Problem	26/05/2021	26/05/2021
							F1905	Device Revision or Replacement	A23	Use of Device Problem	26/05/2021	26/05/2021
							F26	No Health Consequences or Impact	A23	Use of Device Problem	26/05/2021	26/05/2021
					CT2818	Catheterization/Cannulation/Fluid path devices	E2403	No Clinical Signs, Symptoms or Conditions	A23	Use of Device Problem	26/05/2021	26/05/2021
							F1905	Device Revision or Replacement	A23	Use of Device Problem	26/05/2021	26/05/2021
							F26	No Health Consequences or Impact	A23	Use of Device Problem	26/05/2021	26/05/2021
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A23	Use of Device Problem	26/05/2021	26/05/2021
							F1905	Device Revision or Replacement	A23	Use of Device Problem	26/05/2021	26/05/2021
							F26	No Health Consequences or Impact	A23	Use of Device Problem	26/05/2021	26/05/2021
					CT2985	Active	E2403	No Clinical Signs, Symptoms or Conditions	A23	Use of Device Problem	26/05/2021	26/05/2021
							F1905	Device Revision or Replacement	A23	Use of Device Problem	26/05/2021	26/05/2021
							F26	No Health Consequences or Impact	A23	Use of Device Problem	26/05/2021	26/05/2021
					CT3001	Implantable metal	E2403	No Clinical Signs, Symptoms or Conditions	A23	Use of Device Problem	26/05/2021	26/05/2021
							F1905	Device Revision or Replacement	A23	Use of Device Problem	26/05/2021	26/05/2021
							F26	No Health Consequences or Impact	A23	Use of Device Problem	26/05/2021	26/05/2021
					CT3002	Implantable synthetic polymer	E2403	No Clinical Signs, Symptoms or Conditions	A23	Use of Device Problem	26/05/2021	26/05/2021
							F1905	Device Revision or Replacement	A23	Use of Device Problem	26/05/2021	26/05/2021
							F26	No Health Consequences or Impact	A23	Use of Device Problem	26/05/2021	26/05/2021
					CT3006	Implantable inorganic material	E2403	No Clinical Signs, Symptoms or Conditions	A23	Use of Device Problem	26/05/2021	26/05/2021
							F1905	Device Revision or Replacement	A23	Use of Device Problem	26/05/2021	26/05/2021
							F26	No Health Consequences or Impact	A23	Use of Device Problem	26/05/2021	26/05/2021
					CT314	Electrically powered	E2403	No Clinical Signs, Symptoms or Conditions	A23	Use of Device Problem	26/05/2021	26/05/2021
							F1905	Device Revision or Replacement	A23	Use of Device Problem	26/05/2021	26/05/2021
							F26	No Health Consequences or Impact	A23	Use of Device Problem	26/05/2021	26/05/2021
					CT321	Long-term surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A23	Use of Device Problem	26/05/2021	26/05/2021
							F1905	Device Revision or Replacement	A23	Use of Device Problem	26/05/2021	26/05/2021
							F26	No Health Consequences or Impact	A23	Use of Device Problem	26/05/2021	26/05/2021
					CT981	Single use	E2403	No Clinical Signs, Symptoms or Conditions	A23	Use of Device Problem	26/05/2021	26/05/2021
							F1905	Device Revision or Replacement	A23	Use of Device Problem	26/05/2021	26/05/2021
							F26	No Health Consequences or Impact	A23	Use of Device Problem	26/05/2021	26/05/2021
					CT983	Surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A23	Use of Device Problem	26/05/2021	26/05/2021
							F1905	Device Revision or Replacement	A23	Use of Device Problem	26/05/2021	26/05/2021
							F26	No Health Consequences or Impact	A23	Use of Device Problem	26/05/2021	26/05/2021
6000	2021	All other reportable incidents	47270	Cardiac resynchronization therapy implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
6001	2021	All other reportable incidents	17882	Physiologic-monitoring defibrillation system	CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A090812	Missing Test Results		
							F27	Problem Identified During Non-Clinical Procedure	A090812	Missing Test Results		
6002	2022	All other reportable incidents	35223	Endocardial/interventricular septal pacing lead	CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A1204	Failure to Disconnect		
							F1908	Prolonged Surgery	A1204	Failure to Disconnect		
6003	2022	All other reportable incidents	47265	Dual-chamber implantable pacemaker, rate-responsive	CT141	Electrodes/Leads and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A0908	Incorrect, Inadequate or Imprecise Result or Readings	16/02/2022	16/02/2022
							F08	Hospitalization or Prolonged Hospitalization	A0908	Incorrect, Inadequate or Imprecise Result or Readings	16/02/2022	16/02/2022
							F1905	Device Revision or Replacement	A0908	Incorrect, Inadequate or Imprecise Result or Readings	16/02/2022	16/02/2022
					CT145	Cardiovascular devices	E2403	No Clinical Signs, Symptoms or Conditions	A0908	Incorrect, Inadequate or Imprecise Result or Readings	16/02/2022	16/02/2022
							F08	Hospitalization or Prolonged Hospitalization	A0908	Incorrect, Inadequate or Imprecise Result or Readings	16/02/2022	16/02/2022
							F1905	Device Revision or Replacement	A0908	Incorrect, Inadequate or Imprecise Result or Readings	16/02/2022	16/02/2022
					CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A0908	Incorrect, Inadequate or Imprecise Result or Readings	16/02/2022	16/02/2022
							F08	Hospitalization or Prolonged Hospitalization	A0908	Incorrect, Inadequate or Imprecise Result or Readings	16/02/2022	16/02/2022
							F1905	Device Revision or Replacement	A0908	Incorrect, Inadequate or Imprecise Result or Readings	16/02/2022	16/02/2022
					CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A0908	Incorrect, Inadequate or Imprecise Result or Readings	16/02/2022	16/02/2022
							F08	Hospitalization or Prolonged Hospitalization	A0908	Incorrect, Inadequate or Imprecise Result or Readings	16/02/2022	16/02/2022
							F1905	Device Revision or Replacement	A0908	Incorrect, Inadequate or Imprecise Result or Readings	16/02/2022	16/02/2022
					CT233	Surgical	E2403	No Clinical Signs, Symptoms or Conditions	A0908	Incorrect, Inadequate or Imprecise Result or Readings	16/02/2022	16/02/2022
							F08	Hospitalization or Prolonged Hospitalization	A0908	Incorrect, Inadequate or Imprecise Result or Readings	16/02/2022	16/02/2022
							F1905	Device Revision or Replacement	A0908	Incorrect, Inadequate or Imprecise Result or Readings	16/02/2022	16/02/2022

					CT2406	Implantable	E2403	No Clinical Signs, Symptoms or Conditions	A0908	Incorrect, Inadequate or Imprecise Result or Readings	16/02/2022	16/02/2022
							F08	Hospitalization or Prolonged Hospitalization	A0908	Incorrect, Inadequate or Imprecise Result or Readings	16/02/2022	16/02/2022
							F1905	Device Revision or Replacement	A0908	Incorrect, Inadequate or Imprecise Result or Readings	16/02/2022	16/02/2022
					CT2407	Active implantable	E2403	No Clinical Signs, Symptoms or Conditions	A0908	Incorrect, Inadequate or Imprecise Result or Readings	16/02/2022	16/02/2022
							F08	Hospitalization or Prolonged Hospitalization	A0908	Incorrect, Inadequate or Imprecise Result or Readings	16/02/2022	16/02/2022
							F1905	Device Revision or Replacement	A0908	Incorrect, Inadequate or Imprecise Result or Readings	16/02/2022	16/02/2022
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A0908	Incorrect, Inadequate or Imprecise Result or Readings	16/02/2022	16/02/2022
							F08	Hospitalization or Prolonged Hospitalization	A0908	Incorrect, Inadequate or Imprecise Result or Readings	16/02/2022	16/02/2022
							F1905	Device Revision or Replacement	A0908	Incorrect, Inadequate or Imprecise Result or Readings	16/02/2022	16/02/2022
					CT2727	Active implant devices/ancillaries	E2403	No Clinical Signs, Symptoms or Conditions	A0908	Incorrect, Inadequate or Imprecise Result or Readings	16/02/2022	16/02/2022
							F08	Hospitalization or Prolonged Hospitalization	A0908	Incorrect, Inadequate or Imprecise Result or Readings	16/02/2022	16/02/2022
							F1905	Device Revision or Replacement	A0908	Incorrect, Inadequate or Imprecise Result or Readings	16/02/2022	16/02/2022
					CT2750	Implantable cardiac electrical stimulators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A0908	Incorrect, Inadequate or Imprecise Result or Readings	16/02/2022	16/02/2022
							F08	Hospitalization or Prolonged Hospitalization	A0908	Incorrect, Inadequate or Imprecise Result or Readings	16/02/2022	16/02/2022
							F1905	Device Revision or Replacement	A0908	Incorrect, Inadequate or Imprecise Result or Readings	16/02/2022	16/02/2022
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A0908	Incorrect, Inadequate or Imprecise Result or Readings	16/02/2022	16/02/2022
							F08	Hospitalization or Prolonged Hospitalization	A0908	Incorrect, Inadequate or Imprecise Result or Readings	16/02/2022	16/02/2022
							F1905	Device Revision or Replacement	A0908	Incorrect, Inadequate or Imprecise Result or Readings	16/02/2022	16/02/2022
					CT2985	Active	E2403	No Clinical Signs, Symptoms or Conditions	A0908	Incorrect, Inadequate or Imprecise Result or Readings	16/02/2022	16/02/2022
							F08	Hospitalization or Prolonged Hospitalization	A0908	Incorrect, Inadequate or Imprecise Result or Readings	16/02/2022	16/02/2022
							F1905	Device Revision or Replacement	A0908	Incorrect, Inadequate or Imprecise Result or Readings	16/02/2022	16/02/2022
					CT3001	Implantable metal	E2403	No Clinical Signs, Symptoms or Conditions	A0908	Incorrect, Inadequate or Imprecise Result or Readings	16/02/2022	16/02/2022
							F08	Hospitalization or Prolonged Hospitalization	A0908	Incorrect, Inadequate or Imprecise Result or Readings	16/02/2022	16/02/2022
							F1905	Device Revision or Replacement	A0908	Incorrect, Inadequate or Imprecise Result or Readings	16/02/2022	16/02/2022
					CT3002	Implantable synthetic polymer	E2403	No Clinical Signs, Symptoms or Conditions	A0908	Incorrect, Inadequate or Imprecise Result or Readings	16/02/2022	16/02/2022
							F08	Hospitalization or Prolonged Hospitalization	A0908	Incorrect, Inadequate or Imprecise Result or Readings	16/02/2022	16/02/2022
							F1905	Device Revision or Replacement	A0908	Incorrect, Inadequate or Imprecise Result or Readings	16/02/2022	16/02/2022
					CT3006	Implantable inorganic material	E2403	No Clinical Signs, Symptoms or Conditions	A0908	Incorrect, Inadequate or Imprecise Result or Readings	16/02/2022	16/02/2022
							F08	Hospitalization or Prolonged Hospitalization	A0908	Incorrect, Inadequate or Imprecise Result or Readings	16/02/2022	16/02/2022
							F1905	Device Revision or Replacement	A0908	Incorrect, Inadequate or Imprecise Result or Readings	16/02/2022	16/02/2022
					CT314	Electrically powered	E2403	No Clinical Signs, Symptoms or Conditions	A0908	Incorrect, Inadequate or Imprecise Result or Readings	16/02/2022	16/02/2022
							F08	Hospitalization or Prolonged Hospitalization	A0908	Incorrect, Inadequate or Imprecise Result or Readings	16/02/2022	16/02/2022
							F1905	Device Revision or Replacement	A0908	Incorrect, Inadequate or Imprecise Result or Readings	16/02/2022	16/02/2022
					CT321	Long-term surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A0908	Incorrect, Inadequate or Imprecise Result or Readings	16/02/2022	16/02/2022
							F08	Hospitalization or Prolonged Hospitalization	A0908	Incorrect, Inadequate or Imprecise Result or Readings	16/02/2022	16/02/2022
							F1905	Device Revision or Replacement	A0908	Incorrect, Inadequate or Imprecise Result or Readings	16/02/2022	16/02/2022
					CT449	Generators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A0908	Incorrect, Inadequate or Imprecise Result or Readings	16/02/2022	16/02/2022
							F08	Hospitalization or Prolonged Hospitalization	A0908	Incorrect, Inadequate or Imprecise Result or Readings	16/02/2022	16/02/2022
							F1905	Device Revision or Replacement	A0908	Incorrect, Inadequate or Imprecise Result or Readings	16/02/2022	16/02/2022
					CT981	Single use	E2403	No Clinical Signs, Symptoms or Conditions	A0908	Incorrect, Inadequate or Imprecise Result or Readings	16/02/2022	16/02/2022
							F08	Hospitalization or Prolonged Hospitalization	A0908	Incorrect, Inadequate or Imprecise Result or Readings	16/02/2022	16/02/2022
							F1905	Device Revision or Replacement	A0908	Incorrect, Inadequate or Imprecise Result or Readings	16/02/2022	16/02/2022
					CT983	Surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A0908	Incorrect, Inadequate or Imprecise Result or Readings	16/02/2022	16/02/2022
							F08	Hospitalization or Prolonged Hospitalization	A0908	Incorrect, Inadequate or Imprecise Result or Readings	16/02/2022	16/02/2022
							F1905	Device Revision or Replacement	A0908	Incorrect, Inadequate or Imprecise Result or Readings	16/02/2022	16/02/2022
6004	2022	All other reportable incidents	47267	Single-chamber implantable pacemaker, rate-responsive	CT141	Electrodes/Leads and associated devices	E011903	Syncope/Fainting	A070909	Oversensing		
									A0712	Pacing Problem		
							E2403	No Clinical Signs, Symptoms or Conditions	A070909	Oversensing		
									A0712	Pacing Problem		
							F26	No Health Consequences or Impact	A070909	Oversensing		
									A0712	Pacing Problem		
					CT145	Cardiovascular devices	E011903	Syncope/Fainting	A070909	Oversensing		
									A0712	Pacing Problem		

		E2403	No Clinical Signs, Symptoms or Conditions	A070909	Oversensing		
				A0712	Pacing Problem		
		F26	No Health Consequences or Impact	A070909	Oversensing		
				A0712	Pacing Problem		
CT1684	Cardiac pacemakers and associated devices	E011903	Syncope/Fainting	A070909	Oversensing		
				A0712	Pacing Problem		
		E2403	No Clinical Signs, Symptoms or Conditions	A070909	Oversensing		
				A0712	Pacing Problem		
		F26	No Health Consequences or Impact	A070909	Oversensing		
				A0712	Pacing Problem		
CT1685	Cardiac pacemakers	E011903	Syncope/Fainting	A070909	Oversensing		
				A0712	Pacing Problem		
		E2403	No Clinical Signs, Symptoms or Conditions	A070909	Oversensing		
				A0712	Pacing Problem		
		F26	No Health Consequences or Impact	A070909	Oversensing		
				A0712	Pacing Problem		
CT233	Surgical	E011903	Syncope/Fainting	A070909	Oversensing		
				A0712	Pacing Problem		
		E2403	No Clinical Signs, Symptoms or Conditions	A070909	Oversensing		
				A0712	Pacing Problem		
		F26	No Health Consequences or Impact	A070909	Oversensing		
				A0712	Pacing Problem		
CT2406	Implantable	E011903	Syncope/Fainting	A070909	Oversensing		
				A0712	Pacing Problem		
		E2403	No Clinical Signs, Symptoms or Conditions	A070909	Oversensing		
				A0712	Pacing Problem		
		F26	No Health Consequences or Impact	A070909	Oversensing		
				A0712	Pacing Problem		
CT2407	Active implantable	E011903	Syncope/Fainting	A070909	Oversensing		
				A0712	Pacing Problem		
		E2403	No Clinical Signs, Symptoms or Conditions	A070909	Oversensing		
				A0712	Pacing Problem		
		F26	No Health Consequences or Impact	A070909	Oversensing		
				A0712	Pacing Problem		
CT257	Implantable cardiac pacemakers	E011903	Syncope/Fainting	A070909	Oversensing		
				A0712	Pacing Problem		
		E2403	No Clinical Signs, Symptoms or Conditions	A070909	Oversensing		
				A0712	Pacing Problem		
		F26	No Health Consequences or Impact	A070909	Oversensing		
				A0712	Pacing Problem		
CT2727	Active implant devices/ancillaries	E011903	Syncope/Fainting	A070909	Oversensing		
				A0712	Pacing Problem		
		E2403	No Clinical Signs, Symptoms or Conditions	A070909	Oversensing		
				A0712	Pacing Problem		
		F26	No Health Consequences or Impact	A070909	Oversensing		
				A0712	Pacing Problem		
CT2750	Implantable cardiac electrical stimulators and associated devices	E011903	Syncope/Fainting	A070909	Oversensing		
				A0712	Pacing Problem		
		E2403	No Clinical Signs, Symptoms or Conditions	A070909	Oversensing		
				A0712	Pacing Problem		
		F26	No Health Consequences or Impact	A070909	Oversensing		
				A0712	Pacing Problem		
CT2866	Implantable cardiac electrical stimulators	E011903	Syncope/Fainting	A070909	Oversensing		
				A0712	Pacing Problem		
		E2403	No Clinical Signs, Symptoms or Conditions	A070909	Oversensing		
				A0712	Pacing Problem		
		F26	No Health Consequences or Impact	A070909	Oversensing		
				A0712	Pacing Problem		
CT2985	Active	E011903	Syncope/Fainting	A070909	Oversensing		
				A0712	Pacing Problem		
		E2403	No Clinical Signs, Symptoms or Conditions	A070909	Oversensing		
				A0712	Pacing Problem		
		F26	No Health Consequences or Impact	A070909	Oversensing		
				A0712	Pacing Problem		
CT3001	Implantable metal	E011903	Syncope/Fainting	A070909	Oversensing		
				A0712	Pacing Problem		
		E2403	No Clinical Signs, Symptoms or Conditions	A070909	Oversensing		
				A0712	Pacing Problem		
		F26	No Health Consequences or Impact	A070909	Oversensing		
				A0712	Pacing Problem		
CT3002	Implantable synthetic polymer	E011903	Syncope/Fainting	A070909	Oversensing		
				A0712	Pacing Problem		
		E2403	No Clinical Signs, Symptoms or Conditions	A070909	Oversensing		
				A0712	Pacing Problem		
		F26	No Health Consequences or Impact	A070909	Oversensing		
				A0712	Pacing Problem		
CT3006	Implantable inorganic material	E011903	Syncope/Fainting	A070909	Oversensing		
				A0712	Pacing Problem		
		E2403	No Clinical Signs, Symptoms or Conditions	A070909	Oversensing		
				A0712	Pacing Problem		
		F26	No Health Consequences or Impact	A070909	Oversensing		
				A0712	Pacing Problem		
CT314	Electrically powered	E011903	Syncope/Fainting	A070909	Oversensing		
				A0712	Pacing Problem		
		E2403	No Clinical Signs, Symptoms or Conditions	A070909	Oversensing		
				A0712	Pacing Problem		
		F26	No Health Consequences or Impact	A070909	Oversensing		
				A0712	Pacing Problem		

					CT321	Long-term surgical invasive	E011903	Syncope/Fainting	A070909	Oversensing		
									A0712	Pacing Problem		
							E2403	No Clinical Signs, Symptoms or Conditions	A070909	Oversensing		
									A0712	Pacing Problem		
							F26	No Health Consequences or Impact	A070909	Oversensing		
									A0712	Pacing Problem		
					CT449	Generators and associated devices	E011903	Syncope/Fainting	A070909	Oversensing		
									A0712	Pacing Problem		
							E2403	No Clinical Signs, Symptoms or Conditions	A070909	Oversensing		
									A0712	Pacing Problem		
							F26	No Health Consequences or Impact	A070909	Oversensing		
									A0712	Pacing Problem		
					CT981	Single use	E011903	Syncope/Fainting	A070909	Oversensing		
									A0712	Pacing Problem		
							E2403	No Clinical Signs, Symptoms or Conditions	A070909	Oversensing		
									A0712	Pacing Problem		
							F26	No Health Consequences or Impact	A070909	Oversensing		
									A0712	Pacing Problem		
					CT983	Surgical invasive	E011903	Syncope/Fainting	A070909	Oversensing		
									A0712	Pacing Problem		
							E2403	No Clinical Signs, Symptoms or Conditions	A070909	Oversensing		
									A0712	Pacing Problem		
							F26	No Health Consequences or Impact	A070909	Oversensing		
									A0712	Pacing Problem		
6005	2022	All other reportable incidents	47265	Dual-chamber implantable pacemaker, rate-responsive	CT141	Electrodes/Leads and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	01/06/2022	01/06/2022
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	01/06/2022	01/06/2022
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	01/06/2022	01/06/2022
					CT145	Cardiovascular devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	01/06/2022	01/06/2022
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	01/06/2022	01/06/2022
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	01/06/2022	01/06/2022
					CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	01/06/2022	01/06/2022
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	01/06/2022	01/06/2022
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	01/06/2022	01/06/2022
					CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	01/06/2022	01/06/2022
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	01/06/2022	01/06/2022
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	01/06/2022	01/06/2022
					CT233	Surgical	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	01/06/2022	01/06/2022
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	01/06/2022	01/06/2022
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	01/06/2022	01/06/2022
					CT2406	Implantable	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	01/06/2022	01/06/2022
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	01/06/2022	01/06/2022
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	01/06/2022	01/06/2022
					CT2407	Active implantable	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	01/06/2022	01/06/2022
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	01/06/2022	01/06/2022
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	01/06/2022	01/06/2022
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	01/06/2022	01/06/2022
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	01/06/2022	01/06/2022
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	01/06/2022	01/06/2022
					CT2727	Active implant devices/ancillaries	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	01/06/2022	01/06/2022
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	01/06/2022	01/06/2022
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	01/06/2022	01/06/2022
					CT2750	Implantable cardiac electrical stimulators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	01/06/2022	01/06/2022
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	01/06/2022	01/06/2022
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	01/06/2022	01/06/2022
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	01/06/2022	01/06/2022
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	01/06/2022	01/06/2022
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	01/06/2022	01/06/2022
					CT2985	Active	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	01/06/2022	01/06/2022
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	01/06/2022	01/06/2022
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	01/06/2022	01/06/2022
					CT3001	Implantable metal	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	01/06/2022	01/06/2022
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	01/06/2022	01/06/2022
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	01/06/2022	01/06/2022
					CT3002	Implantable synthetic polymer	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	01/06/2022	01/06/2022
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	01/06/2022	01/06/2022
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	01/06/2022	01/06/2022
					CT3006	Implantable inorganic material	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	01/06/2022	01/06/2022
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	01/06/2022	01/06/2022
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	01/06/2022	01/06/2022
					CT314	Electrically powered	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	01/06/2022	01/06/2022
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	01/06/2022	01/06/2022
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	01/06/2022	01/06/2022
					CT321	Long-term surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	01/06/2022	01/06/2022
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	01/06/2022	01/06/2022
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	01/06/2022	01/06/2022
					CT449	Generators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	01/06/2022	01/06/2022
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	01/06/2022	01/06/2022
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	01/06/2022	01/06/2022
					CT981	Single use	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	01/06/2022	01/06/2022
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	01/06/2022	01/06/2022
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	01/06/2022	01/06/2022
					CT983	Surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	01/06/2022	01/06/2022
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	01/06/2022	01/06/2022
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	01/06/2022	01/06/2022
6006	2021	All other reportable incidents	47265	Dual-chamber implantable pacemaker, rate-responsive	CT141	Electrodes/Leads and associated devices	E011903	Syncope/Fainting	A0908	Incorrect, Inadequate or Imprecise Result or Readings		
							E0611	Heart Failure/Congestive Heart Failure	A0908	Incorrect, Inadequate or Imprecise Result or Readings		

							E0623	Unspecified Heart Problem	A0908	Incorrect, Inadequate or Imprecise Result or Readings		
							F08	Hospitalization or Prolonged Hospitalization	A0908	Incorrect, Inadequate or Imprecise Result or Readings		
							F12	Serious Injury/ Illness/ Impairment	A0908	Incorrect, Inadequate or Imprecise Result or Readings		
							F1905	Device Revision or Replacement	A0908	Incorrect, Inadequate or Imprecise Result or Readings		
							F21	Unexpected Deterioration	A0908	Incorrect, Inadequate or Imprecise Result or Readings		
					CT145	Cardiovascular devices	E011903	Syncope/Fainting	A0908	Incorrect, Inadequate or Imprecise Result or Readings		
							E0611	Heart Failure/Congestive Heart Failure	A0908	Incorrect, Inadequate or Imprecise Result or Readings		
							E0623	Unspecified Heart Problem	A0908	Incorrect, Inadequate or Imprecise Result or Readings		
							F08	Hospitalization or Prolonged Hospitalization	A0908	Incorrect, Inadequate or Imprecise Result or Readings		
							F12	Serious Injury/ Illness/ Impairment	A0908	Incorrect, Inadequate or Imprecise Result or Readings		
							F1905	Device Revision or Replacement	A0908	Incorrect, Inadequate or Imprecise Result or Readings		
							F21	Unexpected Deterioration	A0908	Incorrect, Inadequate or Imprecise Result or Readings		
					CT1684	Cardiac pacemakers and associated devices	E011903	Syncope/Fainting	A0908	Incorrect, Inadequate or Imprecise Result or Readings		
							E0611	Heart Failure/Congestive Heart Failure	A0908	Incorrect, Inadequate or Imprecise Result or Readings		
							E0623	Unspecified Heart Problem	A0908	Incorrect, Inadequate or Imprecise Result or Readings		
							F08	Hospitalization or Prolonged Hospitalization	A0908	Incorrect, Inadequate or Imprecise Result or Readings		
							F12	Serious Injury/ Illness/ Impairment	A0908	Incorrect, Inadequate or Imprecise Result or Readings		
							F1905	Device Revision or Replacement	A0908	Incorrect, Inadequate or Imprecise Result or Readings		
							F21	Unexpected Deterioration	A0908	Incorrect, Inadequate or Imprecise Result or Readings		
					CT1685	Cardiac pacemakers	E011903	Syncope/Fainting	A0908	Incorrect, Inadequate or Imprecise Result or Readings		
							E0611	Heart Failure/Congestive Heart Failure	A0908	Incorrect, Inadequate or Imprecise Result or Readings		
							E0623	Unspecified Heart Problem	A0908	Incorrect, Inadequate or Imprecise Result or Readings		
							F08	Hospitalization or Prolonged Hospitalization	A0908	Incorrect, Inadequate or Imprecise Result or Readings		
							F12	Serious Injury/ Illness/ Impairment	A0908	Incorrect, Inadequate or Imprecise Result or Readings		
							F1905	Device Revision or Replacement	A0908	Incorrect, Inadequate or Imprecise Result or Readings		
							F21	Unexpected Deterioration	A0908	Incorrect, Inadequate or Imprecise Result or Readings		
					CT233	Surgical	E011903	Syncope/Fainting	A0908	Incorrect, Inadequate or Imprecise Result or Readings		
							E0611	Heart Failure/Congestive Heart Failure	A0908	Incorrect, Inadequate or Imprecise Result or Readings		
							E0623	Unspecified Heart Problem	A0908	Incorrect, Inadequate or Imprecise Result or Readings		
							F08	Hospitalization or Prolonged Hospitalization	A0908	Incorrect, Inadequate or Imprecise Result or Readings		
							F12	Serious Injury/ Illness/ Impairment	A0908	Incorrect, Inadequate or Imprecise Result or Readings		
							F1905	Device Revision or Replacement	A0908	Incorrect, Inadequate or Imprecise Result or Readings		
							F21	Unexpected Deterioration	A0908	Incorrect, Inadequate or Imprecise Result or Readings		
					CT2406	Implantable	E011903	Syncope/Fainting	A0908	Incorrect, Inadequate or Imprecise Result or Readings		
							E0611	Heart Failure/Congestive Heart Failure	A0908	Incorrect, Inadequate or Imprecise Result or Readings		
							E0623	Unspecified Heart Problem	A0908	Incorrect, Inadequate or Imprecise Result or Readings		
							F08	Hospitalization or Prolonged Hospitalization	A0908	Incorrect, Inadequate or Imprecise Result or Readings		
							F12	Serious Injury/ Illness/ Impairment	A0908	Incorrect, Inadequate or Imprecise Result or Readings		
							F1905	Device Revision or Replacement	A0908	Incorrect, Inadequate or Imprecise Result or Readings		
							F21	Unexpected Deterioration	A0908	Incorrect, Inadequate or Imprecise Result or Readings		
					CT2407	Active implantable	E011903	Syncope/Fainting	A0908	Incorrect, Inadequate or Imprecise Result or Readings		
							E0611	Heart Failure/Congestive Heart Failure	A0908	Incorrect, Inadequate or Imprecise Result or Readings		
							E0623	Unspecified Heart Problem	A0908	Incorrect, Inadequate or Imprecise Result or Readings		
							F08	Hospitalization or Prolonged Hospitalization	A0908	Incorrect, Inadequate or Imprecise Result or Readings		
							F12	Serious Injury/ Illness/ Impairment	A0908	Incorrect, Inadequate or Imprecise Result or Readings		
							F1905	Device Revision or Replacement	A0908	Incorrect, Inadequate or Imprecise Result or Readings		
							F21	Unexpected Deterioration	A0908	Incorrect, Inadequate or Imprecise Result or Readings		
					CT257	Implantable cardiac pacemakers	E011903	Syncope/Fainting	A0908	Incorrect, Inadequate or Imprecise Result or Readings		
							E0611	Heart Failure/Congestive Heart Failure	A0908	Incorrect, Inadequate or Imprecise Result or Readings		
							E0623	Unspecified Heart Problem	A0908	Incorrect, Inadequate or Imprecise Result or Readings		

							F08	Hospitalization or Prolonged Hospitalization	A0908	Incorrect, Inadequate or Imprecise Result or Readings		
							F12	Serious Injury/ Illness/ Impairment	A0908	Incorrect, Inadequate or Imprecise Result or Readings		
							F1905	Device Revision or Replacement	A0908	Incorrect, Inadequate or Imprecise Result or Readings		
							F21	Unexpected Deterioration	A0908	Incorrect, Inadequate or Imprecise Result or Readings		
					CT2727	Active implant devices/ancillaries	E011903	Syncope/Fainting	A0908	Incorrect, Inadequate or Imprecise Result or Readings		
							E0611	Heart Failure/Congestive Heart Failure	A0908	Incorrect, Inadequate or Imprecise Result or Readings		
							E0623	Unspecified Heart Problem	A0908	Incorrect, Inadequate or Imprecise Result or Readings		
							F08	Hospitalization or Prolonged Hospitalization	A0908	Incorrect, Inadequate or Imprecise Result or Readings		
							F12	Serious Injury/ Illness/ Impairment	A0908	Incorrect, Inadequate or Imprecise Result or Readings		
							F1905	Device Revision or Replacement	A0908	Incorrect, Inadequate or Imprecise Result or Readings		
							F21	Unexpected Deterioration	A0908	Incorrect, Inadequate or Imprecise Result or Readings		
					CT2750	Implantable cardiac electrical stimulators and associated devices	E011903	Syncope/Fainting	A0908	Incorrect, Inadequate or Imprecise Result or Readings		
							E0611	Heart Failure/Congestive Heart Failure	A0908	Incorrect, Inadequate or Imprecise Result or Readings		
							E0623	Unspecified Heart Problem	A0908	Incorrect, Inadequate or Imprecise Result or Readings		
							F08	Hospitalization or Prolonged Hospitalization	A0908	Incorrect, Inadequate or Imprecise Result or Readings		
							F12	Serious Injury/ Illness/ Impairment	A0908	Incorrect, Inadequate or Imprecise Result or Readings		
							F1905	Device Revision or Replacement	A0908	Incorrect, Inadequate or Imprecise Result or Readings		
							F21	Unexpected Deterioration	A0908	Incorrect, Inadequate or Imprecise Result or Readings		
					CT2866	Implantable cardiac electrical stimulators	E011903	Syncope/Fainting	A0908	Incorrect, Inadequate or Imprecise Result or Readings		
							E0611	Heart Failure/Congestive Heart Failure	A0908	Incorrect, Inadequate or Imprecise Result or Readings		
							E0623	Unspecified Heart Problem	A0908	Incorrect, Inadequate or Imprecise Result or Readings		
							F08	Hospitalization or Prolonged Hospitalization	A0908	Incorrect, Inadequate or Imprecise Result or Readings		
							F12	Serious Injury/ Illness/ Impairment	A0908	Incorrect, Inadequate or Imprecise Result or Readings		
							F1905	Device Revision or Replacement	A0908	Incorrect, Inadequate or Imprecise Result or Readings		
							F21	Unexpected Deterioration	A0908	Incorrect, Inadequate or Imprecise Result or Readings		
					CT2985	Active	E011903	Syncope/Fainting	A0908	Incorrect, Inadequate or Imprecise Result or Readings		
							E0611	Heart Failure/Congestive Heart Failure	A0908	Incorrect, Inadequate or Imprecise Result or Readings		
							E0623	Unspecified Heart Problem	A0908	Incorrect, Inadequate or Imprecise Result or Readings		
							F08	Hospitalization or Prolonged Hospitalization	A0908	Incorrect, Inadequate or Imprecise Result or Readings		
							F12	Serious Injury/ Illness/ Impairment	A0908	Incorrect, Inadequate or Imprecise Result or Readings		
							F1905	Device Revision or Replacement	A0908	Incorrect, Inadequate or Imprecise Result or Readings		
							F21	Unexpected Deterioration	A0908	Incorrect, Inadequate or Imprecise Result or Readings		
					CT3001	Implantable metal	E011903	Syncope/Fainting	A0908	Incorrect, Inadequate or Imprecise Result or Readings		
							E0611	Heart Failure/Congestive Heart Failure	A0908	Incorrect, Inadequate or Imprecise Result or Readings		
							E0623	Unspecified Heart Problem	A0908	Incorrect, Inadequate or Imprecise Result or Readings		
							F08	Hospitalization or Prolonged Hospitalization	A0908	Incorrect, Inadequate or Imprecise Result or Readings		
							F12	Serious Injury/ Illness/ Impairment	A0908	Incorrect, Inadequate or Imprecise Result or Readings		
							F1905	Device Revision or Replacement	A0908	Incorrect, Inadequate or Imprecise Result or Readings		
							F21	Unexpected Deterioration	A0908	Incorrect, Inadequate or Imprecise Result or Readings		
					CT3002	Implantable synthetic polymer	E011903	Syncope/Fainting	A0908	Incorrect, Inadequate or Imprecise Result or Readings		
							E0611	Heart Failure/Congestive Heart Failure	A0908	Incorrect, Inadequate or Imprecise Result or Readings		
							E0623	Unspecified Heart Problem	A0908	Incorrect, Inadequate or Imprecise Result or Readings		
							F08	Hospitalization or Prolonged Hospitalization	A0908	Incorrect, Inadequate or Imprecise Result or Readings		
							F12	Serious Injury/ Illness/ Impairment	A0908	Incorrect, Inadequate or Imprecise Result or Readings		
							F1905	Device Revision or Replacement	A0908	Incorrect, Inadequate or Imprecise Result or Readings		
							F21	Unexpected Deterioration	A0908	Incorrect, Inadequate or Imprecise Result or Readings		
					CT3006	Implantable inorganic material	E011903	Syncope/Fainting	A0908	Incorrect, Inadequate or Imprecise Result or Readings		
							E0611	Heart Failure/Congestive Heart Failure	A0908	Incorrect, Inadequate or Imprecise Result or Readings		
							E0623	Unspecified Heart Problem	A0908	Incorrect, Inadequate or Imprecise Result or Readings		
							F08	Hospitalization or Prolonged Hospitalization	A0908	Incorrect, Inadequate or Imprecise Result or Readings		

							F12	Serious Injury/ Illness/ Impairment	A0908	Incorrect, Inadequate or Imprecise Result or Readings		
							F1905	Device Revision or Replacement	A0908	Incorrect, Inadequate or Imprecise Result or Readings		
							F21	Unexpected Deterioration	A0908	Incorrect, Inadequate or Imprecise Result or Readings		
					CT314	Electrically powered	E011903	Syncope/Fainting	A0908	Incorrect, Inadequate or Imprecise Result or Readings		
							E0611	Heart Failure/Congestive Heart Failure	A0908	Incorrect, Inadequate or Imprecise Result or Readings		
							E0623	Unspecified Heart Problem	A0908	Incorrect, Inadequate or Imprecise Result or Readings		
							F08	Hospitalization or Prolonged Hospitalization	A0908	Incorrect, Inadequate or Imprecise Result or Readings		
							F12	Serious Injury/ Illness/ Impairment	A0908	Incorrect, Inadequate or Imprecise Result or Readings		
							F1905	Device Revision or Replacement	A0908	Incorrect, Inadequate or Imprecise Result or Readings		
							F21	Unexpected Deterioration	A0908	Incorrect, Inadequate or Imprecise Result or Readings		
					CT321	Long-term surgical invasive	E011903	Syncope/Fainting	A0908	Incorrect, Inadequate or Imprecise Result or Readings		
							E0611	Heart Failure/Congestive Heart Failure	A0908	Incorrect, Inadequate or Imprecise Result or Readings		
							E0623	Unspecified Heart Problem	A0908	Incorrect, Inadequate or Imprecise Result or Readings		
							F08	Hospitalization or Prolonged Hospitalization	A0908	Incorrect, Inadequate or Imprecise Result or Readings		
							F12	Serious Injury/ Illness/ Impairment	A0908	Incorrect, Inadequate or Imprecise Result or Readings		
							F1905	Device Revision or Replacement	A0908	Incorrect, Inadequate or Imprecise Result or Readings		
							F21	Unexpected Deterioration	A0908	Incorrect, Inadequate or Imprecise Result or Readings		
					CT449	Generators and associated devices	E011903	Syncope/Fainting	A0908	Incorrect, Inadequate or Imprecise Result or Readings		
							E0611	Heart Failure/Congestive Heart Failure	A0908	Incorrect, Inadequate or Imprecise Result or Readings		
							E0623	Unspecified Heart Problem	A0908	Incorrect, Inadequate or Imprecise Result or Readings		
							F08	Hospitalization or Prolonged Hospitalization	A0908	Incorrect, Inadequate or Imprecise Result or Readings		
							F12	Serious Injury/ Illness/ Impairment	A0908	Incorrect, Inadequate or Imprecise Result or Readings		
							F1905	Device Revision or Replacement	A0908	Incorrect, Inadequate or Imprecise Result or Readings		
							F21	Unexpected Deterioration	A0908	Incorrect, Inadequate or Imprecise Result or Readings		
					CT981	Single use	E011903	Syncope/Fainting	A0908	Incorrect, Inadequate or Imprecise Result or Readings		
							E0611	Heart Failure/Congestive Heart Failure	A0908	Incorrect, Inadequate or Imprecise Result or Readings		
							E0623	Unspecified Heart Problem	A0908	Incorrect, Inadequate or Imprecise Result or Readings		
							F08	Hospitalization or Prolonged Hospitalization	A0908	Incorrect, Inadequate or Imprecise Result or Readings		
							F12	Serious Injury/ Illness/ Impairment	A0908	Incorrect, Inadequate or Imprecise Result or Readings		
							F1905	Device Revision or Replacement	A0908	Incorrect, Inadequate or Imprecise Result or Readings		
							F21	Unexpected Deterioration	A0908	Incorrect, Inadequate or Imprecise Result or Readings		
					CT983	Surgical invasive	E011903	Syncope/Fainting	A0908	Incorrect, Inadequate or Imprecise Result or Readings		
							E0611	Heart Failure/Congestive Heart Failure	A0908	Incorrect, Inadequate or Imprecise Result or Readings		
							E0623	Unspecified Heart Problem	A0908	Incorrect, Inadequate or Imprecise Result or Readings		
							F08	Hospitalization or Prolonged Hospitalization	A0908	Incorrect, Inadequate or Imprecise Result or Readings		
							F12	Serious Injury/ Illness/ Impairment	A0908	Incorrect, Inadequate or Imprecise Result or Readings		
							F1905	Device Revision or Replacement	A0908	Incorrect, Inadequate or Imprecise Result or Readings		
							F21	Unexpected Deterioration	A0908	Incorrect, Inadequate or Imprecise Result or Readings		
6007	2021	All other reportable incidents	36241	Pacemaker/defibrillator lead	CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A072201	High Impedance		
							F26	No Health Consequences or Impact	A072201	High Impedance		
6008	2021	All other reportable incidents	47265	Dual-chamber implantable pacemaker, rate-responsive	CT141	Electrodes/Leads and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	18/06/2021	18/06/2021
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	18/06/2021	18/06/2021
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	18/06/2021	18/06/2021
					CT145	Cardiovascular devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	18/06/2021	18/06/2021
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	18/06/2021	18/06/2021
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	18/06/2021	18/06/2021
					CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	18/06/2021	18/06/2021
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	18/06/2021	18/06/2021
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	18/06/2021	18/06/2021
					CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	18/06/2021	18/06/2021
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	18/06/2021	18/06/2021
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	18/06/2021	18/06/2021
					CT233	Surgical	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	18/06/2021	18/06/2021
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	18/06/2021	18/06/2021
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	18/06/2021	18/06/2021
					CT2406	Implantable	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	18/06/2021	18/06/2021
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	18/06/2021	18/06/2021
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	18/06/2021	18/06/2021
					CT2407	Active implantable	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	18/06/2021	18/06/2021

						F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	18/06/2021	18/06/2021
						F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	18/06/2021	18/06/2021
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	18/06/2021	18/06/2021
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	18/06/2021	18/06/2021
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	18/06/2021	18/06/2021
					CT2727	Active implant devices/ancillaries	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	18/06/2021	18/06/2021
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	18/06/2021	18/06/2021
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	18/06/2021	18/06/2021
					CT2750	Implantable cardiac electrical stimulators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	18/06/2021	18/06/2021
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	18/06/2021	18/06/2021
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	18/06/2021	18/06/2021
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	18/06/2021	18/06/2021
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	18/06/2021	18/06/2021
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	18/06/2021	18/06/2021
					CT2985	Active	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	18/06/2021	18/06/2021
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	18/06/2021	18/06/2021
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	18/06/2021	18/06/2021
					CT3001	Implantable metal	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	18/06/2021	18/06/2021
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	18/06/2021	18/06/2021
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	18/06/2021	18/06/2021
					CT3002	Implantable synthetic polymer	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	18/06/2021	18/06/2021
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	18/06/2021	18/06/2021
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	18/06/2021	18/06/2021
					CT3006	Implantable inorganic material	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	18/06/2021	18/06/2021
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	18/06/2021	18/06/2021
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	18/06/2021	18/06/2021
					CT314	Electrically powered	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	18/06/2021	18/06/2021
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	18/06/2021	18/06/2021
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	18/06/2021	18/06/2021
					CT321	Long-term surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	18/06/2021	18/06/2021
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	18/06/2021	18/06/2021
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	18/06/2021	18/06/2021
					CT449	Generators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	18/06/2021	18/06/2021
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	18/06/2021	18/06/2021
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	18/06/2021	18/06/2021
					CT981	Single use	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	18/06/2021	18/06/2021
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	18/06/2021	18/06/2021
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	18/06/2021	18/06/2021
					CT983	Surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	18/06/2021	18/06/2021
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	18/06/2021	18/06/2021
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	18/06/2021	18/06/2021
6009	2022	All other reportable incidents	35852	Single-chamber implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem	07/03/2022	07/03/2022
									A070504	Premature Discharge of Battery	07/03/2022	07/03/2022
							F08	Hospitalization or Prolonged Hospitalization	A0705	Battery Problem	07/03/2022	07/03/2022
									A070504	Premature Discharge of Battery	07/03/2022	07/03/2022
							F1905	Device Revision or Replacement	A0705	Battery Problem	07/03/2022	07/03/2022
									A070504	Premature Discharge of Battery	07/03/2022	07/03/2022
6010	2021	All other reportable incidents	47265	Dual-chamber implantable pacemaker, rate-responsive	CT141	Electrodes/Leads and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	19/08/2021	19/08/2021
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	19/08/2021	19/08/2021
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	19/08/2021	19/08/2021
					CT145	Cardiovascular devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	19/08/2021	19/08/2021
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	19/08/2021	19/08/2021
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	19/08/2021	19/08/2021
					CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	19/08/2021	19/08/2021
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	19/08/2021	19/08/2021
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	19/08/2021	19/08/2021
					CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	19/08/2021	19/08/2021
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	19/08/2021	19/08/2021
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	19/08/2021	19/08/2021
					CT233	Surgical	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	19/08/2021	19/08/2021
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	19/08/2021	19/08/2021
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	19/08/2021	19/08/2021
					CT2406	Implantable	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	19/08/2021	19/08/2021
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	19/08/2021	19/08/2021
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	19/08/2021	19/08/2021
					CT2407	Active implantable	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	19/08/2021	19/08/2021
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	19/08/2021	19/08/2021
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	19/08/2021	19/08/2021
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	19/08/2021	19/08/2021
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	19/08/2021	19/08/2021
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	19/08/2021	19/08/2021
					CT2727	Active implant devices/ancillaries	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	19/08/2021	19/08/2021
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	19/08/2021	19/08/2021
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	19/08/2021	19/08/2021
					CT2750	Implantable cardiac electrical stimulators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	19/08/2021	19/08/2021
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	19/08/2021	19/08/2021
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	19/08/2021	19/08/2021
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	19/08/2021	19/08/2021
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	19/08/2021	19/08/2021
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	19/08/2021	19/08/2021
					CT2985	Active	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	19/08/2021	19/08/2021
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	19/08/2021	19/08/2021
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	19/08/2021	19/08/2021
					CT3001	Implantable metal	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	19/08/2021	19/08/2021
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	19/08/2021	19/08/2021
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	19/08/2021	19/08/2021
					CT3002	Implantable synthetic polymer	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	19/08/2021	19/08/2021
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	19/08/2021	19/08/2021

						F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	19/08/2021	19/08/2021
				CT3006	Implantable inorganic material	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	19/08/2021	19/08/2021
						F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	19/08/2021	19/08/2021
						F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	19/08/2021	19/08/2021
				CT314	Electrically powered	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	19/08/2021	19/08/2021
						F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	19/08/2021	19/08/2021
						F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	19/08/2021	19/08/2021
				CT321	Long-term surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	19/08/2021	19/08/2021
						F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	19/08/2021	19/08/2021
						F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	19/08/2021	19/08/2021
				CT449	Generators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	19/08/2021	19/08/2021
						F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	19/08/2021	19/08/2021
						F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	19/08/2021	19/08/2021
				CT981	Single use	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	19/08/2021	19/08/2021
						F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	19/08/2021	19/08/2021
						F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	19/08/2021	19/08/2021
				CT983	Surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	19/08/2021	19/08/2021
						F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	19/08/2021	19/08/2021
						F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	19/08/2021	19/08/2021
6011	2021	All other reportable incidents	47265	Dual-chamber implantable pacemaker, rate-responsive	CT141	Electrodes/Leads and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	21/04/2021	21/04/2021
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	21/04/2021	21/04/2021
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	21/04/2021	21/04/2021
					CT145	Cardiovascular devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	21/04/2021	21/04/2021
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	21/04/2021	21/04/2021
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	21/04/2021	21/04/2021
					CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	21/04/2021	21/04/2021
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	21/04/2021	21/04/2021
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	21/04/2021	21/04/2021
					CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	21/04/2021	21/04/2021
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	21/04/2021	21/04/2021
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	21/04/2021	21/04/2021
					CT233	Surgical	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	21/04/2021	21/04/2021
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	21/04/2021	21/04/2021
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	21/04/2021	21/04/2021
					CT2406	Implantable	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	21/04/2021	21/04/2021
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	21/04/2021	21/04/2021
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	21/04/2021	21/04/2021
					CT2407	Active implantable	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	21/04/2021	21/04/2021
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	21/04/2021	21/04/2021
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	21/04/2021	21/04/2021
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	21/04/2021	21/04/2021
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	21/04/2021	21/04/2021
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	21/04/2021	21/04/2021
					CT2727	Active implant devices/ancillaries	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	21/04/2021	21/04/2021
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	21/04/2021	21/04/2021
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	21/04/2021	21/04/2021
					CT2750	Implantable cardiac electrical stimulators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	21/04/2021	21/04/2021
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	21/04/2021	21/04/2021
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	21/04/2021	21/04/2021
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	21/04/2021	21/04/2021
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	21/04/2021	21/04/2021
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	21/04/2021	21/04/2021
					CT2985	Active	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	21/04/2021	21/04/2021
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	21/04/2021	21/04/2021
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	21/04/2021	21/04/2021
					CT3001	Implantable metal	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	21/04/2021	21/04/2021
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	21/04/2021	21/04/2021
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	21/04/2021	21/04/2021
					CT3002	Implantable synthetic polymer	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	21/04/2021	21/04/2021
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	21/04/2021	21/04/2021
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	21/04/2021	21/04/2021
					CT3006	Implantable inorganic material	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	21/04/2021	21/04/2021
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	21/04/2021	21/04/2021
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	21/04/2021	21/04/2021
					CT314	Electrically powered	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	21/04/2021	21/04/2021
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	21/04/2021	21/04/2021
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	21/04/2021	21/04/2021
					CT321	Long-term surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	21/04/2021	21/04/2021
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	21/04/2021	21/04/2021
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	21/04/2021	21/04/2021
					CT449	Generators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	21/04/2021	21/04/2021
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	21/04/2021	21/04/2021
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	21/04/2021	21/04/2021
					CT981	Single use	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	21/04/2021	21/04/2021
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	21/04/2021	21/04/2021
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	21/04/2021	21/04/2021
					CT983	Surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	21/04/2021	21/04/2021
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	21/04/2021	21/04/2021
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	21/04/2021	21/04/2021
6012	2021	All other reportable incidents	47265	Dual-chamber implantable pacemaker, rate-responsive	CT141	Electrodes/Leads and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	16/07/2021	16/07/2021
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	16/07/2021	16/07/2021
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	16/07/2021	16/07/2021
					CT145	Cardiovascular devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	16/07/2021	16/07/2021
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	16/07/2021	16/07/2021
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	16/07/2021	16/07/2021
					CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	16/07/2021	16/07/2021
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	16/07/2021	16/07/2021
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	16/07/2021	16/07/2021

					CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	16/07/2021	16/07/2021
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	16/07/2021	16/07/2021
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	16/07/2021	16/07/2021
					CT233	Surgical	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	16/07/2021	16/07/2021
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	16/07/2021	16/07/2021
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	16/07/2021	16/07/2021
					CT2406	Implantable	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	16/07/2021	16/07/2021
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	16/07/2021	16/07/2021
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	16/07/2021	16/07/2021
					CT2407	Active implantable	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	16/07/2021	16/07/2021
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	16/07/2021	16/07/2021
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	16/07/2021	16/07/2021
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	16/07/2021	16/07/2021
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	16/07/2021	16/07/2021
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	16/07/2021	16/07/2021
					CT2727	Active implant devices/ancillaries	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	16/07/2021	16/07/2021
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	16/07/2021	16/07/2021
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	16/07/2021	16/07/2021
					CT2750	Implantable cardiac electrical stimulators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	16/07/2021	16/07/2021
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	16/07/2021	16/07/2021
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	16/07/2021	16/07/2021
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	16/07/2021	16/07/2021
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	16/07/2021	16/07/2021
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	16/07/2021	16/07/2021
					CT2985	Active	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	16/07/2021	16/07/2021
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	16/07/2021	16/07/2021
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	16/07/2021	16/07/2021
					CT3001	Implantable metal	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	16/07/2021	16/07/2021
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	16/07/2021	16/07/2021
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	16/07/2021	16/07/2021
					CT3002	Implantable synthetic polymer	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	16/07/2021	16/07/2021
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	16/07/2021	16/07/2021
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	16/07/2021	16/07/2021
					CT3006	Implantable inorganic material	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	16/07/2021	16/07/2021
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	16/07/2021	16/07/2021
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	16/07/2021	16/07/2021
					CT314	Electrically powered	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	16/07/2021	16/07/2021
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	16/07/2021	16/07/2021
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	16/07/2021	16/07/2021
					CT321	Long-term surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	16/07/2021	16/07/2021
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	16/07/2021	16/07/2021
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	16/07/2021	16/07/2021
					CT449	Generators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	16/07/2021	16/07/2021
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	16/07/2021	16/07/2021
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	16/07/2021	16/07/2021
					CT981	Single use	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	16/07/2021	16/07/2021
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	16/07/2021	16/07/2021
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	16/07/2021	16/07/2021
					CT983	Surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	16/07/2021	16/07/2021
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	16/07/2021	16/07/2021
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	16/07/2021	16/07/2021
6013	2021	All other reportable incidents	36241	Pacemaker/defibrillator lead	CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070909	Oversensing	25/06/2021	25/06/2021
							F08	Hospitalization or Prolonged Hospitalization	A070909	Oversensing	25/06/2021	25/06/2021
							F1905	Device Revision or Replacement	A070909	Oversensing	25/06/2021	25/06/2021
6014	2021	All other reportable incidents	35852	Single-chamber implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	05/08/2021	05/08/2021
							F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A070504	Premature Discharge of Battery	05/08/2021	05/08/2021
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	05/08/2021	05/08/2021
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery	05/08/2021	05/08/2021
6015	2021	All other reportable incidents	36241	Pacemaker/defibrillator lead	CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A072201	High Impedance		
							F26	No Health Consequences or Impact	A072201	High Impedance		
6016	2022	All other reportable incidents	47265	Dual-chamber implantable pacemaker, rate-responsive	CT141	Electrodes/Leads and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem	07/01/2022	07/01/2022
									A070504	Premature Discharge of Battery	07/01/2022	07/01/2022
							F08	Hospitalization or Prolonged Hospitalization	A0705	Battery Problem	07/01/2022	07/01/2022
									A070504	Premature Discharge of Battery	07/01/2022	07/01/2022
							F1905	Device Revision or Replacement	A0705	Battery Problem	07/01/2022	07/01/2022
									A070504	Premature Discharge of Battery	07/01/2022	07/01/2022
					CT145	Cardiovascular devices	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem	07/01/2022	07/01/2022
									A070504	Premature Discharge of Battery	07/01/2022	07/01/2022
							F08	Hospitalization or Prolonged Hospitalization	A0705	Battery Problem	07/01/2022	07/01/2022
									A070504	Premature Discharge of Battery	07/01/2022	07/01/2022
							F1905	Device Revision or Replacement	A0705	Battery Problem	07/01/2022	07/01/2022
									A070504	Premature Discharge of Battery	07/01/2022	07/01/2022
					CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem	07/01/2022	07/01/2022
									A070504	Premature Discharge of Battery	07/01/2022	07/01/2022
							F08	Hospitalization or Prolonged Hospitalization	A0705	Battery Problem	07/01/2022	07/01/2022
									A070504	Premature Discharge of Battery	07/01/2022	07/01/2022
							F1905	Device Revision or Replacement	A0705	Battery Problem	07/01/2022	07/01/2022
									A070504	Premature Discharge of Battery	07/01/2022	07/01/2022
					CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem	07/01/2022	07/01/2022
									A070504	Premature Discharge of Battery	07/01/2022	07/01/2022
							F08	Hospitalization or Prolonged Hospitalization	A0705	Battery Problem	07/01/2022	07/01/2022
									A070504	Premature Discharge of Battery	07/01/2022	07/01/2022
							F1905	Device Revision or Replacement	A0705	Battery Problem	07/01/2022	07/01/2022
									A070504	Premature Discharge of Battery	07/01/2022	07/01/2022
					CT233	Surgical	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem	07/01/2022	07/01/2022
									A070504	Premature Discharge of Battery	07/01/2022	07/01/2022
							F08	Hospitalization or Prolonged Hospitalization	A0705	Battery Problem	07/01/2022	07/01/2022
									A070504	Premature Discharge of Battery	07/01/2022	07/01/2022

		F1905	Device Revision or Replacement	A0705	Battery Problem	07/01/2022	07/01/2022
				A070504	Premature Discharge of Battery	07/01/2022	07/01/2022
CT2406	Implantable	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem	07/01/2022	07/01/2022
				A070504	Premature Discharge of Battery	07/01/2022	07/01/2022
		F08	Hospitalization or Prolonged Hospitalization	A0705	Battery Problem	07/01/2022	07/01/2022
				A070504	Premature Discharge of Battery	07/01/2022	07/01/2022
		F1905	Device Revision or Replacement	A0705	Battery Problem	07/01/2022	07/01/2022
				A070504	Premature Discharge of Battery	07/01/2022	07/01/2022
CT2407	Active implantable	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem	07/01/2022	07/01/2022
				A070504	Premature Discharge of Battery	07/01/2022	07/01/2022
		F08	Hospitalization or Prolonged Hospitalization	A0705	Battery Problem	07/01/2022	07/01/2022
				A070504	Premature Discharge of Battery	07/01/2022	07/01/2022
		F1905	Device Revision or Replacement	A0705	Battery Problem	07/01/2022	07/01/2022
				A070504	Premature Discharge of Battery	07/01/2022	07/01/2022
CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem	07/01/2022	07/01/2022
				A070504	Premature Discharge of Battery	07/01/2022	07/01/2022
		F08	Hospitalization or Prolonged Hospitalization	A0705	Battery Problem	07/01/2022	07/01/2022
				A070504	Premature Discharge of Battery	07/01/2022	07/01/2022
		F1905	Device Revision or Replacement	A0705	Battery Problem	07/01/2022	07/01/2022
				A070504	Premature Discharge of Battery	07/01/2022	07/01/2022
CT2727	Active implant devices/ancillaries	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem	07/01/2022	07/01/2022
				A070504	Premature Discharge of Battery	07/01/2022	07/01/2022
		F08	Hospitalization or Prolonged Hospitalization	A0705	Battery Problem	07/01/2022	07/01/2022
				A070504	Premature Discharge of Battery	07/01/2022	07/01/2022
		F1905	Device Revision or Replacement	A0705	Battery Problem	07/01/2022	07/01/2022
				A070504	Premature Discharge of Battery	07/01/2022	07/01/2022
CT2750	Implantable cardiac electrical stimulators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem	07/01/2022	07/01/2022
				A070504	Premature Discharge of Battery	07/01/2022	07/01/2022
		F08	Hospitalization or Prolonged Hospitalization	A0705	Battery Problem	07/01/2022	07/01/2022
				A070504	Premature Discharge of Battery	07/01/2022	07/01/2022
		F1905	Device Revision or Replacement	A0705	Battery Problem	07/01/2022	07/01/2022
				A070504	Premature Discharge of Battery	07/01/2022	07/01/2022
CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem	07/01/2022	07/01/2022
				A070504	Premature Discharge of Battery	07/01/2022	07/01/2022
		F08	Hospitalization or Prolonged Hospitalization	A0705	Battery Problem	07/01/2022	07/01/2022
				A070504	Premature Discharge of Battery	07/01/2022	07/01/2022
		F1905	Device Revision or Replacement	A0705	Battery Problem	07/01/2022	07/01/2022
				A070504	Premature Discharge of Battery	07/01/2022	07/01/2022
CT2985	Active	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem	07/01/2022	07/01/2022
				A070504	Premature Discharge of Battery	07/01/2022	07/01/2022
		F08	Hospitalization or Prolonged Hospitalization	A0705	Battery Problem	07/01/2022	07/01/2022
				A070504	Premature Discharge of Battery	07/01/2022	07/01/2022
		F1905	Device Revision or Replacement	A0705	Battery Problem	07/01/2022	07/01/2022
				A070504	Premature Discharge of Battery	07/01/2022	07/01/2022
CT3001	Implantable metal	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem	07/01/2022	07/01/2022
				A070504	Premature Discharge of Battery	07/01/2022	07/01/2022
		F08	Hospitalization or Prolonged Hospitalization	A0705	Battery Problem	07/01/2022	07/01/2022
				A070504	Premature Discharge of Battery	07/01/2022	07/01/2022
		F1905	Device Revision or Replacement	A0705	Battery Problem	07/01/2022	07/01/2022
				A070504	Premature Discharge of Battery	07/01/2022	07/01/2022
CT3002	Implantable synthetic polymer	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem	07/01/2022	07/01/2022
				A070504	Premature Discharge of Battery	07/01/2022	07/01/2022
		F08	Hospitalization or Prolonged Hospitalization	A0705	Battery Problem	07/01/2022	07/01/2022
				A070504	Premature Discharge of Battery	07/01/2022	07/01/2022
		F1905	Device Revision or Replacement	A0705	Battery Problem	07/01/2022	07/01/2022
				A070504	Premature Discharge of Battery	07/01/2022	07/01/2022
CT3006	Implantable inorganic material	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem	07/01/2022	07/01/2022
				A070504	Premature Discharge of Battery	07/01/2022	07/01/2022
		F08	Hospitalization or Prolonged Hospitalization	A0705	Battery Problem	07/01/2022	07/01/2022
				A070504	Premature Discharge of Battery	07/01/2022	07/01/2022
		F1905	Device Revision or Replacement	A0705	Battery Problem	07/01/2022	07/01/2022
				A070504	Premature Discharge of Battery	07/01/2022	07/01/2022
CT314	Electrically powered	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem	07/01/2022	07/01/2022
				A070504	Premature Discharge of Battery	07/01/2022	07/01/2022
		F08	Hospitalization or Prolonged Hospitalization	A0705	Battery Problem	07/01/2022	07/01/2022
				A070504	Premature Discharge of Battery	07/01/2022	07/01/2022
		F1905	Device Revision or Replacement	A0705	Battery Problem	07/01/2022	07/01/2022
				A070504	Premature Discharge of Battery	07/01/2022	07/01/2022
CT321	Long-term surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem	07/01/2022	07/01/2022
				A070504	Premature Discharge of Battery	07/01/2022	07/01/2022
		F08	Hospitalization or Prolonged Hospitalization	A0705	Battery Problem	07/01/2022	07/01/2022
				A070504	Premature Discharge of Battery	07/01/2022	07/01/2022
		F1905	Device Revision or Replacement	A0705	Battery Problem	07/01/2022	07/01/2022
				A070504	Premature Discharge of Battery	07/01/2022	07/01/2022
CT449	Generators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem	07/01/2022	07/01/2022
				A070504	Premature Discharge of Battery	07/01/2022	07/01/2022
		F08	Hospitalization or Prolonged Hospitalization	A0705	Battery Problem	07/01/2022	07/01/2022
				A070504	Premature Discharge of Battery	07/01/2022	07/01/2022
		F1905	Device Revision or Replacement	A0705	Battery Problem	07/01/2022	07/01/2022
				A070504	Premature Discharge of Battery	07/01/2022	07/01/2022
CT981	Single use	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem	07/01/2022	07/01/2022
				A070504	Premature Discharge of Battery	07/01/2022	07/01/2022
		F08	Hospitalization or Prolonged Hospitalization	A0705	Battery Problem	07/01/2022	07/01/2022
				A070504	Premature Discharge of Battery	07/01/2022	07/01/2022
		F1905	Device Revision or Replacement	A0705	Battery Problem	07/01/2022	07/01/2022
				A070504	Premature Discharge of Battery	07/01/2022	07/01/2022
CT983	Surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem	07/01/2022	07/01/2022
				A070504	Premature Discharge of Battery	07/01/2022	07/01/2022

							F08	Hospitalization or Prolonged Hospitalization	A0705	Battery Problem	07/01/2022	07/01/2022
									A070504	Premature Discharge of Battery	07/01/2022	07/01/2022
							F1905	Device Revision or Replacement	A0705	Battery Problem	07/01/2022	07/01/2022
									A070504	Premature Discharge of Battery	07/01/2022	07/01/2022
6017	2022	All other reportable incidents	36241	Pacemaker/defibrillator lead	CT1684	Cardiac pacemakers and associated devices	E0604	Cardiac Perforation	A051201	Device Dislodged or Dislocated	04/01/2022	04/01/2022
									A070101	Failure to Capture	04/01/2022	04/01/2022
									A072202	Low Impedance	04/01/2022	04/01/2022
									A1301	Failure to Read Input Signal	04/01/2022	04/01/2022
							E0619	Pericardial Effusion	A051201	Device Dislodged or Dislocated	04/01/2022	04/01/2022
									A070101	Failure to Capture	04/01/2022	04/01/2022
									A072202	Low Impedance	04/01/2022	04/01/2022
									A1301	Failure to Read Input Signal	04/01/2022	04/01/2022
							E0717	Dyspnea	A051201	Device Dislodged or Dislocated	04/01/2022	04/01/2022
									A070101	Failure to Capture	04/01/2022	04/01/2022
									A072202	Low Impedance	04/01/2022	04/01/2022
									A1301	Failure to Read Input Signal	04/01/2022	04/01/2022
							E1020	Nausea	A051201	Device Dislodged or Dislocated	04/01/2022	04/01/2022
									A070101	Failure to Capture	04/01/2022	04/01/2022
									A072202	Low Impedance	04/01/2022	04/01/2022
									A1301	Failure to Read Input Signal	04/01/2022	04/01/2022
							E2403	No Clinical Signs, Symptoms or Conditions	A051201	Device Dislodged or Dislocated	04/01/2022	04/01/2022
									A070101	Failure to Capture	04/01/2022	04/01/2022
									A072202	Low Impedance	04/01/2022	04/01/2022
									A1301	Failure to Read Input Signal	04/01/2022	04/01/2022
							F08	Hospitalization or Prolonged Hospitalization	A051201	Device Dislodged or Dislocated	04/01/2022	04/01/2022
									A070101	Failure to Capture	04/01/2022	04/01/2022
									A072202	Low Impedance	04/01/2022	04/01/2022
									A1301	Failure to Read Input Signal	04/01/2022	04/01/2022
							F19	Surgical Intervention	A051201	Device Dislodged or Dislocated	04/01/2022	04/01/2022
									A070101	Failure to Capture	04/01/2022	04/01/2022
									A072202	Low Impedance	04/01/2022	04/01/2022
									A1301	Failure to Read Input Signal	04/01/2022	04/01/2022
							F1905	Device Revision or Replacement	A051201	Device Dislodged or Dislocated	04/01/2022	04/01/2022
									A070101	Failure to Capture	04/01/2022	04/01/2022
									A072202	Low Impedance	04/01/2022	04/01/2022
									A1301	Failure to Read Input Signal	04/01/2022	04/01/2022
							F2203	Imaging Required	A051201	Device Dislodged or Dislocated	04/01/2022	04/01/2022
									A070101	Failure to Capture	04/01/2022	04/01/2022
									A072202	Low Impedance	04/01/2022	04/01/2022
									A1301	Failure to Read Input Signal	04/01/2022	04/01/2022
6018	2022	All other reportable incidents	17882	Physiologic-monitoring defibrillation system	CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070908	Failure to Sense		
							F26	No Health Consequences or Impact	A070908	Failure to Sense		
6019	2021	All other reportable incidents	35039	Epicardial pacing lead	CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070909	Oversensing		
6020	2021	All other reportable incidents	35852	Single-chamber implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem	21/07/2021	21/07/2021
									A070504	Premature Discharge of Battery	21/07/2021	21/07/2021
							F08	Hospitalization or Prolonged Hospitalization	A0705	Battery Problem	21/07/2021	21/07/2021
									A070504	Premature Discharge of Battery	21/07/2021	21/07/2021
							F1905	Device Revision or Replacement	A0705	Battery Problem	21/07/2021	21/07/2021
									A070504	Premature Discharge of Battery	21/07/2021	21/07/2021
6021	2021	All other reportable incidents	47265	Dual-chamber implantable pacemaker, rate-responsive	CT141	Electrodes/Leads and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	06/07/2021	06/07/2021
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery	06/07/2021	06/07/2021
					CT145	Cardiovascular devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	06/07/2021	06/07/2021
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery	06/07/2021	06/07/2021
					CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	06/07/2021	06/07/2021
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery	06/07/2021	06/07/2021
					CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	06/07/2021	06/07/2021
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery	06/07/2021	06/07/2021
					CT233	Surgical	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	06/07/2021	06/07/2021
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery	06/07/2021	06/07/2021
					CT2406	Implantable	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	06/07/2021	06/07/2021
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery	06/07/2021	06/07/2021
					CT2407	Active implantable	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	06/07/2021	06/07/2021
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery	06/07/2021	06/07/2021
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	06/07/2021	06/07/2021
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery	06/07/2021	06/07/2021
					CT2727	Active implant devices/ancillaries	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	06/07/2021	06/07/2021
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery	06/07/2021	06/07/2021
					CT2750	Implantable cardiac electrical stimulators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	06/07/2021	06/07/2021
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery	06/07/2021	06/07/2021
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	06/07/2021	06/07/2021
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery	06/07/2021	06/07/2021
					CT2985	Active	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	06/07/2021	06/07/2021
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery	06/07/2021	06/07/2021
					CT3001	Implantable metal	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	06/07/2021	06/07/2021
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery	06/07/2021	06/07/2021
					CT3002	Implantable synthetic polymer	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	06/07/2021	06/07/2021
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery	06/07/2021	06/07/2021
					CT3006	Implantable inorganic material	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	06/07/2021	06/07/2021
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery	06/07/2021	06/07/2021
					CT314	Electrically powered	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	06/07/2021	06/07/2021
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery	06/07/2021	06/07/2021
					CT321	Long-term surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	06/07/2021	06/07/2021
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery	06/07/2021	06/07/2021
					CT449	Generators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	06/07/2021	06/07/2021
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery	06/07/2021	06/07/2021
					CT981	Single use	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	06/07/2021	06/07/2021
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery	06/07/2021	06/07/2021
					CT983	Surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	06/07/2021	06/07/2021

							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery	06/07/2021	06/07/2021
6022	2021	All other reportable incidents	35223	Endocardial/interventricular septal pacing lead	CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A07	Electrical /Electronic Property Problem		
									A0722	Impedance Problem		
									A072201	High Impedance		
							F26	No Health Consequences or Impact	A07	Electrical /Electronic Property Problem		
									A0722	Impedance Problem		
									A072201	High Impedance		
6023	2021	All other reportable incidents	35852	Single-chamber implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E2104	Electric Shock	A070909	Oversensing		
									A071302	Inappropriate/Inadequate Shock/Stimulation		
							F13	Misdiagnosis/ Misclassification	A070909	Oversensing		
									A071302	Inappropriate/Inadequate Shock/Stimulation		
6024	2021	All other reportable incidents	47265	Dual-chamber implantable pacemaker, rate-responsive	CT141	Electrodes/Leads and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A0709	Device Sensing Problem		
									A09	Output Problem		
							F26	No Health Consequences or Impact	A0709	Device Sensing Problem		
									A09	Output Problem		
					CT145	Cardiovascular devices	E2403	No Clinical Signs, Symptoms or Conditions	A0709	Device Sensing Problem		
									A09	Output Problem		
							F26	No Health Consequences or Impact	A0709	Device Sensing Problem		
									A09	Output Problem		
					CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A0709	Device Sensing Problem		
									A09	Output Problem		
							F26	No Health Consequences or Impact	A0709	Device Sensing Problem		
									A09	Output Problem		
					CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A0709	Device Sensing Problem		
									A09	Output Problem		
							F26	No Health Consequences or Impact	A0709	Device Sensing Problem		
									A09	Output Problem		
					CT233	Surgical	E2403	No Clinical Signs, Symptoms or Conditions	A0709	Device Sensing Problem		
									A09	Output Problem		
							F26	No Health Consequences or Impact	A0709	Device Sensing Problem		
									A09	Output Problem		
					CT2406	Implantable	E2403	No Clinical Signs, Symptoms or Conditions	A0709	Device Sensing Problem		
									A09	Output Problem		
							F26	No Health Consequences or Impact	A0709	Device Sensing Problem		
									A09	Output Problem		
					CT2407	Active implantable	E2403	No Clinical Signs, Symptoms or Conditions	A0709	Device Sensing Problem		
									A09	Output Problem		
							F26	No Health Consequences or Impact	A0709	Device Sensing Problem		
									A09	Output Problem		
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A0709	Device Sensing Problem		
									A09	Output Problem		
							F26	No Health Consequences or Impact	A0709	Device Sensing Problem		
									A09	Output Problem		
					CT2727	Active implant devices/ancillaries	E2403	No Clinical Signs, Symptoms or Conditions	A0709	Device Sensing Problem		
									A09	Output Problem		
							F26	No Health Consequences or Impact	A0709	Device Sensing Problem		
									A09	Output Problem		
					CT2750	Implantable cardiac electrical stimulators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A0709	Device Sensing Problem		
									A09	Output Problem		
							F26	No Health Consequences or Impact	A0709	Device Sensing Problem		
									A09	Output Problem		
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A0709	Device Sensing Problem		
									A09	Output Problem		
							F26	No Health Consequences or Impact	A0709	Device Sensing Problem		
									A09	Output Problem		
					CT2985	Active	E2403	No Clinical Signs, Symptoms or Conditions	A0709	Device Sensing Problem		
									A09	Output Problem		
							F26	No Health Consequences or Impact	A0709	Device Sensing Problem		
									A09	Output Problem		
					CT3001	Implantable metal	E2403	No Clinical Signs, Symptoms or Conditions	A0709	Device Sensing Problem		
									A09	Output Problem		
							F26	No Health Consequences or Impact	A0709	Device Sensing Problem		
									A09	Output Problem		
					CT3002	Implantable synthetic polymer	E2403	No Clinical Signs, Symptoms or Conditions	A0709	Device Sensing Problem		
									A09	Output Problem		
							F26	No Health Consequences or Impact	A0709	Device Sensing Problem		
									A09	Output Problem		
					CT3006	Implantable inorganic material	E2403	No Clinical Signs, Symptoms or Conditions	A0709	Device Sensing Problem		
									A09	Output Problem		
							F26	No Health Consequences or Impact	A0709	Device Sensing Problem		
									A09	Output Problem		
					CT314	Electrically powered	E2403	No Clinical Signs, Symptoms or Conditions	A0709	Device Sensing Problem		
									A09	Output Problem		
							F26	No Health Consequences or Impact	A0709	Device Sensing Problem		
									A09	Output Problem		
					CT321	Long-term surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A0709	Device Sensing Problem		
									A09	Output Problem		
							F26	No Health Consequences or Impact	A0709	Device Sensing Problem		
									A09	Output Problem		
					CT449	Generators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A0709	Device Sensing Problem		
									A09	Output Problem		
							F26	No Health Consequences or Impact	A0709	Device Sensing Problem		
									A09	Output Problem		
					CT981	Single use	E2403	No Clinical Signs, Symptoms or Conditions	A0709	Device Sensing Problem		
									A09	Output Problem		
							F26	No Health Consequences or Impact	A0709	Device Sensing Problem		
									A09	Output Problem		
					CT983	Surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A0709	Device Sensing Problem		

									A09	Output Problem		
						F26	No Health Consequences or Impact	A0709	Device Sensing Problem			
								A09	Output Problem			
6025	2022	All other reportable incidents	47263	Cardiac resynchronization therapy implantable pacemaker	CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A0908	Incorrect, Inadequate or Imprecise Result or Readings	27/01/2022	27/01/2022
							F08	Hospitalization or Prolonged Hospitalization	A0908	Incorrect, Inadequate or Imprecise Result or Readings	27/01/2022	27/01/2022
							F1905	Device Revision or Replacement	A0908	Incorrect, Inadequate or Imprecise Result or Readings	27/01/2022	27/01/2022
					CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A0908	Incorrect, Inadequate or Imprecise Result or Readings	27/01/2022	27/01/2022
							F08	Hospitalization or Prolonged Hospitalization	A0908	Incorrect, Inadequate or Imprecise Result or Readings	27/01/2022	27/01/2022
							F1905	Device Revision or Replacement	A0908	Incorrect, Inadequate or Imprecise Result or Readings	27/01/2022	27/01/2022
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A0908	Incorrect, Inadequate or Imprecise Result or Readings	27/01/2022	27/01/2022
							F08	Hospitalization or Prolonged Hospitalization	A0908	Incorrect, Inadequate or Imprecise Result or Readings	27/01/2022	27/01/2022
							F1905	Device Revision or Replacement	A0908	Incorrect, Inadequate or Imprecise Result or Readings	27/01/2022	27/01/2022
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A0908	Incorrect, Inadequate or Imprecise Result or Readings	27/01/2022	27/01/2022
							F08	Hospitalization or Prolonged Hospitalization	A0908	Incorrect, Inadequate or Imprecise Result or Readings	27/01/2022	27/01/2022
							F1905	Device Revision or Replacement	A0908	Incorrect, Inadequate or Imprecise Result or Readings	27/01/2022	27/01/2022
6026	2022	All other reportable incidents	47267	Single-chamber implantable pacemaker, rate-responsive	CT141	Electrodes/Leads and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070101	Failure to Capture		
							F1905	Device Revision or Replacement	A070101	Failure to Capture		
					CT145	Cardiovascular devices	E2403	No Clinical Signs, Symptoms or Conditions	A070101	Failure to Capture		
							F1905	Device Revision or Replacement	A070101	Failure to Capture		
					CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070101	Failure to Capture		
							F1905	Device Revision or Replacement	A070101	Failure to Capture		
					CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A070101	Failure to Capture		
							F1905	Device Revision or Replacement	A070101	Failure to Capture		
					CT233	Surgical	E2403	No Clinical Signs, Symptoms or Conditions	A070101	Failure to Capture		
							F1905	Device Revision or Replacement	A070101	Failure to Capture		
					CT2406	Implantable	E2403	No Clinical Signs, Symptoms or Conditions	A070101	Failure to Capture		
							F1905	Device Revision or Replacement	A070101	Failure to Capture		
					CT2407	Active implantable	E2403	No Clinical Signs, Symptoms or Conditions	A070101	Failure to Capture		
							F1905	Device Revision or Replacement	A070101	Failure to Capture		
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A070101	Failure to Capture		
							F1905	Device Revision or Replacement	A070101	Failure to Capture		
					CT2727	Active implant devices/ancillaries	E2403	No Clinical Signs, Symptoms or Conditions	A070101	Failure to Capture		
							F1905	Device Revision or Replacement	A070101	Failure to Capture		
					CT2750	Implantable cardiac electrical stimulators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070101	Failure to Capture		
							F1905	Device Revision or Replacement	A070101	Failure to Capture		
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A070101	Failure to Capture		
							F1905	Device Revision or Replacement	A070101	Failure to Capture		
					CT2985	Active	E2403	No Clinical Signs, Symptoms or Conditions	A070101	Failure to Capture		
							F1905	Device Revision or Replacement	A070101	Failure to Capture		
					CT3001	Implantable metal	E2403	No Clinical Signs, Symptoms or Conditions	A070101	Failure to Capture		
							F1905	Device Revision or Replacement	A070101	Failure to Capture		
					CT3002	Implantable synthetic polymer	E2403	No Clinical Signs, Symptoms or Conditions	A070101	Failure to Capture		
							F1905	Device Revision or Replacement	A070101	Failure to Capture		
					CT3006	Implantable inorganic material	E2403	No Clinical Signs, Symptoms or Conditions	A070101	Failure to Capture		
							F1905	Device Revision or Replacement	A070101	Failure to Capture		
					CT314	Electrically powered	E2403	No Clinical Signs, Symptoms or Conditions	A070101	Failure to Capture		
							F1905	Device Revision or Replacement	A070101	Failure to Capture		
					CT321	Long-term surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A070101	Failure to Capture		
							F1905	Device Revision or Replacement	A070101	Failure to Capture		
					CT449	Generators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070101	Failure to Capture		
							F1905	Device Revision or Replacement	A070101	Failure to Capture		
					CT981	Single use	E2403	No Clinical Signs, Symptoms or Conditions	A070101	Failure to Capture		
							F1905	Device Revision or Replacement	A070101	Failure to Capture		
					CT983	Surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A070101	Failure to Capture		
							F1905	Device Revision or Replacement	A070101	Failure to Capture		
6027	2022	All other reportable incidents	47265	Dual-chamber implantable pacemaker, rate-responsive	CT141	Electrodes/Leads and associated devices	E2402	Appropriate Clinical Signs, Symptoms and Conditions Term/Code not Available	A070101	Failure to Capture		
									A090811	Unable to Obtain Readings		
							E2403	No Clinical Signs, Symptoms or Conditions	A070101	Failure to Capture		
									A090811	Unable to Obtain Readings		
							F28	Appropriate Health Impact Term/Code not Available	A070101	Failure to Capture		
									A090811	Unable to Obtain Readings		
					CT145	Cardiovascular devices	E2402	Appropriate Clinical Signs, Symptoms and Conditions Term/Code not Available	A070101	Failure to Capture		
									A090811	Unable to Obtain Readings		
							E2403	No Clinical Signs, Symptoms or Conditions	A070101	Failure to Capture		
									A090811	Unable to Obtain Readings		
							F28	Appropriate Health Impact Term/Code not Available	A070101	Failure to Capture		
									A090811	Unable to Obtain Readings		
					CT1684	Cardiac pacemakers and associated devices	E2402	Appropriate Clinical Signs, Symptoms and Conditions Term/Code not Available	A070101	Failure to Capture		
									A090811	Unable to Obtain Readings		
							E2403	No Clinical Signs, Symptoms or Conditions	A070101	Failure to Capture		
									A090811	Unable to Obtain Readings		
							F28	Appropriate Health Impact Term/Code not Available	A070101	Failure to Capture		
									A090811	Unable to Obtain Readings		
					CT1685	Cardiac pacemakers	E2402	Appropriate Clinical Signs, Symptoms and Conditions Term/Code not Available	A070101	Failure to Capture		
									A090811	Unable to Obtain Readings		
							E2403	No Clinical Signs, Symptoms or Conditions	A070101	Failure to Capture		
									A090811	Unable to Obtain Readings		
							F28	Appropriate Health Impact Term/Code not Available	A070101	Failure to Capture		
									A090811	Unable to Obtain Readings		

					CT233	Surgical	E2402	Appropriate Clinical Signs, Symptoms and Conditions Term/Code not Available	A070101	Failure to Capture
									A090811	Unable to Obtain Readings
							E2403	No Clinical Signs, Symptoms or Conditions	A070101	Failure to Capture
									A090811	Unable to Obtain Readings
							F28	Appropriate Health Impact Term/Code not Available	A070101	Failure to Capture
									A090811	Unable to Obtain Readings
					CT2406	Implantable	E2402	Appropriate Clinical Signs, Symptoms and Conditions Term/Code not Available	A070101	Failure to Capture
									A090811	Unable to Obtain Readings
							E2403	No Clinical Signs, Symptoms or Conditions	A070101	Failure to Capture
									A090811	Unable to Obtain Readings
							F28	Appropriate Health Impact Term/Code not Available	A070101	Failure to Capture
									A090811	Unable to Obtain Readings
					CT2407	Active implantable	E2402	Appropriate Clinical Signs, Symptoms and Conditions Term/Code not Available	A070101	Failure to Capture
									A090811	Unable to Obtain Readings
							E2403	No Clinical Signs, Symptoms or Conditions	A070101	Failure to Capture
									A090811	Unable to Obtain Readings
							F28	Appropriate Health Impact Term/Code not Available	A070101	Failure to Capture
									A090811	Unable to Obtain Readings
					CT257	Implantable cardiac pacemakers	E2402	Appropriate Clinical Signs, Symptoms and Conditions Term/Code not Available	A070101	Failure to Capture
									A090811	Unable to Obtain Readings
							E2403	No Clinical Signs, Symptoms or Conditions	A070101	Failure to Capture
									A090811	Unable to Obtain Readings
							F28	Appropriate Health Impact Term/Code not Available	A070101	Failure to Capture
									A090811	Unable to Obtain Readings
					CT2727	Active implant devices/ancillaries	E2402	Appropriate Clinical Signs, Symptoms and Conditions Term/Code not Available	A070101	Failure to Capture
									A090811	Unable to Obtain Readings
							E2403	No Clinical Signs, Symptoms or Conditions	A070101	Failure to Capture
									A090811	Unable to Obtain Readings
							F28	Appropriate Health Impact Term/Code not Available	A070101	Failure to Capture
									A090811	Unable to Obtain Readings
					CT2750	Implantable cardiac electrical stimulators and associated devices	E2402	Appropriate Clinical Signs, Symptoms and Conditions Term/Code not Available	A070101	Failure to Capture
									A090811	Unable to Obtain Readings
							E2403	No Clinical Signs, Symptoms or Conditions	A070101	Failure to Capture
									A090811	Unable to Obtain Readings
							F28	Appropriate Health Impact Term/Code not Available	A070101	Failure to Capture
									A090811	Unable to Obtain Readings
					CT2866	Implantable cardiac electrical stimulators	E2402	Appropriate Clinical Signs, Symptoms and Conditions Term/Code not Available	A070101	Failure to Capture
									A090811	Unable to Obtain Readings
							E2403	No Clinical Signs, Symptoms or Conditions	A070101	Failure to Capture
									A090811	Unable to Obtain Readings
							F28	Appropriate Health Impact Term/Code not Available	A070101	Failure to Capture
									A090811	Unable to Obtain Readings
					CT2985	Active	E2402	Appropriate Clinical Signs, Symptoms and Conditions Term/Code not Available	A070101	Failure to Capture
									A090811	Unable to Obtain Readings
							E2403	No Clinical Signs, Symptoms or Conditions	A070101	Failure to Capture
									A090811	Unable to Obtain Readings
							F28	Appropriate Health Impact Term/Code not Available	A070101	Failure to Capture
									A090811	Unable to Obtain Readings
					CT3001	Implantable metal	E2402	Appropriate Clinical Signs, Symptoms and Conditions Term/Code not Available	A070101	Failure to Capture
									A090811	Unable to Obtain Readings
							E2403	No Clinical Signs, Symptoms or Conditions	A070101	Failure to Capture
									A090811	Unable to Obtain Readings
							F28	Appropriate Health Impact Term/Code not Available	A070101	Failure to Capture
									A090811	Unable to Obtain Readings
					CT3002	Implantable synthetic polymer	E2402	Appropriate Clinical Signs, Symptoms and Conditions Term/Code not Available	A070101	Failure to Capture
									A090811	Unable to Obtain Readings
							E2403	No Clinical Signs, Symptoms or Conditions	A070101	Failure to Capture
									A090811	Unable to Obtain Readings
							F28	Appropriate Health Impact Term/Code not Available	A070101	Failure to Capture
									A090811	Unable to Obtain Readings
					CT3006	Implantable inorganic material	E2402	Appropriate Clinical Signs, Symptoms and Conditions Term/Code not Available	A070101	Failure to Capture
									A090811	Unable to Obtain Readings
							E2403	No Clinical Signs, Symptoms or Conditions	A070101	Failure to Capture
									A090811	Unable to Obtain Readings
							F28	Appropriate Health Impact Term/Code not Available	A070101	Failure to Capture
									A090811	Unable to Obtain Readings
					CT314	Electrically powered	E2402	Appropriate Clinical Signs, Symptoms and Conditions Term/Code not Available	A070101	Failure to Capture
									A090811	Unable to Obtain Readings
							E2403	No Clinical Signs, Symptoms or Conditions	A070101	Failure to Capture
									A090811	Unable to Obtain Readings
							F28	Appropriate Health Impact Term/Code not Available	A070101	Failure to Capture
									A090811	Unable to Obtain Readings
					CT321	Long-term surgical invasive	E2402	Appropriate Clinical Signs, Symptoms and Conditions Term/Code not Available	A070101	Failure to Capture
									A090811	Unable to Obtain Readings
							E2403	No Clinical Signs, Symptoms or Conditions	A070101	Failure to Capture
									A090811	Unable to Obtain Readings
							F28	Appropriate Health Impact Term/Code not Available	A070101	Failure to Capture
									A090811	Unable to Obtain Readings
					CT449	Generators and associated devices	E2402	Appropriate Clinical Signs, Symptoms and Conditions Term/Code not Available	A070101	Failure to Capture
									A090811	Unable to Obtain Readings
							E2403	No Clinical Signs, Symptoms or Conditions	A070101	Failure to Capture
									A090811	Unable to Obtain Readings
							F28	Appropriate Health Impact Term/Code not Available	A070101	Failure to Capture
									A090811	Unable to Obtain Readings
					CT981	Single use	E2402	Appropriate Clinical Signs, Symptoms and Conditions Term/Code not Available	A070101	Failure to Capture
									A090811	Unable to Obtain Readings
							E2403	No Clinical Signs, Symptoms or Conditions	A070101	Failure to Capture
									A090811	Unable to Obtain Readings

							F28	Appropriate Health Impact Term/Code not Available	A070101	Failure to Capture		
									A090811	Unable to Obtain Readings		
					CT983	Surgical invasive	E2402	Appropriate Clinical Signs, Symptoms and Conditions Term/Code not Available	A070101	Failure to Capture		
									A090811	Unable to Obtain Readings		
							E2403	No Clinical Signs, Symptoms or Conditions	A070101	Failure to Capture		
									A090811	Unable to Obtain Readings		
							F28	Appropriate Health Impact Term/Code not Available	A070101	Failure to Capture		
									A090811	Unable to Obtain Readings		
6028	2021	All other reportable incidents	35852	Single-chamber implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E2104	Electric Shock	A071302	Inappropriate/Inadequate Shock/Stimulation		
							E2403	No Clinical Signs, Symptoms or Conditions	A071302	Inappropriate/Inadequate Shock/Stimulation		
							F13	Misdiagnosis/ Misclassification	A071302	Inappropriate/Inadequate Shock/Stimulation		
6029	2022	All other reportable incidents	36241	Pacemaker/defibrillator lead	CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A150207	Difficult to Remove	18/11/2021	18/11/2021
							F1908	Prolonged Surgery	A150207	Difficult to Remove	18/11/2021	18/11/2021
6030	2021	All other reportable incidents	47270	Cardiac resynchronization therapy implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E011903	Syncope/Fainting	A070909	Oversensing		
									A071204	No Pacing		
							F21	Unexpected Deterioration	A070909	Oversensing		
									A071204	No Pacing		
							F26	No Health Consequences or Impact	A070909	Oversensing		
									A071204	No Pacing		
6031	2022	All other reportable incidents	36241	Pacemaker/defibrillator lead	CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070102	High Capture Threshold	10/02/2022	10/02/2022
							F08	Hospitalization or Prolonged Hospitalization	A070102	High Capture Threshold	10/02/2022	10/02/2022
							F1905	Device Revision or Replacement	A070102	High Capture Threshold	10/02/2022	10/02/2022
6032	2022	Unanticipated serious deterioration in state of health	47263	Cardiac resynchronization therapy implantable pacemaker	CT1684	Cardiac pacemakers and associated devices	E060109	Tachycardia	A27	Appropriate Device Problem Term/Code not Available	08/02/2022	08/02/2022
							F1203	Life Threatening Illness or Injury	A27	Appropriate Device Problem Term/Code not Available	08/02/2022	08/02/2022
							F1903	Device Explantation	A27	Appropriate Device Problem Term/Code not Available	08/02/2022	08/02/2022
							F23	Unexpected Medical Intervention	A27	Appropriate Device Problem Term/Code not Available	08/02/2022	08/02/2022
					CT1685	Cardiac pacemakers	E060109	Tachycardia	A27	Appropriate Device Problem Term/Code not Available	08/02/2022	08/02/2022
							F1203	Life Threatening Illness or Injury	A27	Appropriate Device Problem Term/Code not Available	08/02/2022	08/02/2022
							F1903	Device Explantation	A27	Appropriate Device Problem Term/Code not Available	08/02/2022	08/02/2022
							F23	Unexpected Medical Intervention	A27	Appropriate Device Problem Term/Code not Available	08/02/2022	08/02/2022
					CT257	Implantable cardiac pacemakers	E060109	Tachycardia	A27	Appropriate Device Problem Term/Code not Available	08/02/2022	08/02/2022
							F1203	Life Threatening Illness or Injury	A27	Appropriate Device Problem Term/Code not Available	08/02/2022	08/02/2022
							F1903	Device Explantation	A27	Appropriate Device Problem Term/Code not Available	08/02/2022	08/02/2022
							F23	Unexpected Medical Intervention	A27	Appropriate Device Problem Term/Code not Available	08/02/2022	08/02/2022
					CT2866	Implantable cardiac electrical stimulators	E060109	Tachycardia	A27	Appropriate Device Problem Term/Code not Available	08/02/2022	08/02/2022
							F1203	Life Threatening Illness or Injury	A27	Appropriate Device Problem Term/Code not Available	08/02/2022	08/02/2022
							F1903	Device Explantation	A27	Appropriate Device Problem Term/Code not Available	08/02/2022	08/02/2022
							F23	Unexpected Medical Intervention	A27	Appropriate Device Problem Term/Code not Available	08/02/2022	08/02/2022
6033	2022	All other reportable incidents	47267	Single-chamber implantable pacemaker, rate-responsive	CT141	Electrodes/Leads and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	15/03/2022	15/03/2022
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	15/03/2022	15/03/2022
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	15/03/2022	15/03/2022
					CT145	Cardiovascular devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	15/03/2022	15/03/2022
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	15/03/2022	15/03/2022
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	15/03/2022	15/03/2022
					CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	15/03/2022	15/03/2022
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	15/03/2022	15/03/2022
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	15/03/2022	15/03/2022
					CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	15/03/2022	15/03/2022
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	15/03/2022	15/03/2022
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	15/03/2022	15/03/2022
					CT233	Surgical	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	15/03/2022	15/03/2022
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	15/03/2022	15/03/2022
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	15/03/2022	15/03/2022
					CT2406	Implantable	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	15/03/2022	15/03/2022
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	15/03/2022	15/03/2022
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	15/03/2022	15/03/2022
					CT2407	Active implantable	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	15/03/2022	15/03/2022
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	15/03/2022	15/03/2022
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	15/03/2022	15/03/2022
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	15/03/2022	15/03/2022
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	15/03/2022	15/03/2022
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	15/03/2022	15/03/2022
					CT2727	Active implant devices/ancillaries	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	15/03/2022	15/03/2022
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	15/03/2022	15/03/2022
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	15/03/2022	15/03/2022
					CT2750	Implantable cardiac electrical stimulators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	15/03/2022	15/03/2022
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	15/03/2022	15/03/2022
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	15/03/2022	15/03/2022
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	15/03/2022	15/03/2022
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	15/03/2022	15/03/2022
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	15/03/2022	15/03/2022
					CT2985	Active	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	15/03/2022	15/03/2022
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	15/03/2022	15/03/2022
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	15/03/2022	15/03/2022
					CT3001	Implantable metal	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	15/03/2022	15/03/2022
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	15/03/2022	15/03/2022

							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	15/03/2022	15/03/2022
					CT3002	Implantable synthetic polymer	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	15/03/2022	15/03/2022
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	15/03/2022	15/03/2022
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	15/03/2022	15/03/2022
					CT3006	Implantable inorganic material	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	15/03/2022	15/03/2022
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	15/03/2022	15/03/2022
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	15/03/2022	15/03/2022
					CT314	Electrically powered	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	15/03/2022	15/03/2022
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	15/03/2022	15/03/2022
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	15/03/2022	15/03/2022
					CT321	Long-term surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	15/03/2022	15/03/2022
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	15/03/2022	15/03/2022
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	15/03/2022	15/03/2022
					CT449	Generators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	15/03/2022	15/03/2022
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	15/03/2022	15/03/2022
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	15/03/2022	15/03/2022
					CT981	Single use	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	15/03/2022	15/03/2022
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	15/03/2022	15/03/2022
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	15/03/2022	15/03/2022
					CT983	Surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	15/03/2022	15/03/2022
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	15/03/2022	15/03/2022
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	15/03/2022	15/03/2022
6034	2021	All other reportable incidents	47270	Cardiac resynchronization therapy implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
6035	2021	All other reportable incidents	47270	Cardiac resynchronization therapy implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
6036	2021	All other reportable incidents	47263	Cardiac resynchronization therapy implantable pacemaker	CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	06/07/2021	06/07/2021
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	06/07/2021	06/07/2021
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery	06/07/2021	06/07/2021
					CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	06/07/2021	06/07/2021
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	06/07/2021	06/07/2021
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery	06/07/2021	06/07/2021
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	06/07/2021	06/07/2021
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	06/07/2021	06/07/2021
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery	06/07/2021	06/07/2021
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	06/07/2021	06/07/2021
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	06/07/2021	06/07/2021
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery	06/07/2021	06/07/2021
6037	2022	All other reportable incidents	35223	Endocardial/interventricular septal pacing lead	CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A0401	Break		
									A070909	Oversensing		
									A090801	Signal Artifact/Noise		
							F26	No Health Consequences or Impact	A0401	Break		
									A070909	Oversensing		
									A090801	Signal Artifact/Noise		
6038	2022	All other reportable incidents	37265	Dual-chamber implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A070909	Oversensing		
							F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A070909	Oversensing		
6039	2022	All other reportable incidents	47270	Cardiac resynchronization therapy implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A072201	High Impedance		
							F26	No Health Consequences or Impact	A072201	High Impedance		
6040	2022	All other reportable incidents	47267	Single-chamber implantable pacemaker, rate-responsive	CT141	Electrodes/Leads and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	27/05/2022	27/05/2022
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	27/05/2022	27/05/2022
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	27/05/2022	27/05/2022
					CT145	Cardiovascular devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	27/05/2022	27/05/2022
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	27/05/2022	27/05/2022
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	27/05/2022	27/05/2022
					CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	27/05/2022	27/05/2022
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	27/05/2022	27/05/2022
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	27/05/2022	27/05/2022
					CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	27/05/2022	27/05/2022
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	27/05/2022	27/05/2022
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	27/05/2022	27/05/2022
					CT233	Surgical	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	27/05/2022	27/05/2022
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	27/05/2022	27/05/2022
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	27/05/2022	27/05/2022
					CT2406	Implantable	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	27/05/2022	27/05/2022
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	27/05/2022	27/05/2022
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	27/05/2022	27/05/2022
					CT2407	Active implantable	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	27/05/2022	27/05/2022
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	27/05/2022	27/05/2022
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	27/05/2022	27/05/2022
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	27/05/2022	27/05/2022
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	27/05/2022	27/05/2022
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	27/05/2022	27/05/2022
					CT2727	Active implant devices/ancillaries	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	27/05/2022	27/05/2022
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	27/05/2022	27/05/2022
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	27/05/2022	27/05/2022
					CT2750	Implantable cardiac electrical stimulators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	27/05/2022	27/05/2022
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	27/05/2022	27/05/2022
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	27/05/2022	27/05/2022
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	27/05/2022	27/05/2022
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	27/05/2022	27/05/2022
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	27/05/2022	27/05/2022
					CT2985	Active	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	27/05/2022	27/05/2022
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	27/05/2022	27/05/2022
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	27/05/2022	27/05/2022
					CT3001	Implantable metal	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	27/05/2022	27/05/2022
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	27/05/2022	27/05/2022
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	27/05/2022	27/05/2022
					CT3002	Implantable synthetic polymer	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	27/05/2022	27/05/2022

						F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	27/05/2022	27/05/2022
						F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	27/05/2022	27/05/2022
					CT3006	Implantable inorganic material	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	27/05/2022	27/05/2022
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	27/05/2022	27/05/2022
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	27/05/2022	27/05/2022
					CT314	Electrically powered	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	27/05/2022	27/05/2022
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	27/05/2022	27/05/2022
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	27/05/2022	27/05/2022
					CT321	Long-term surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	27/05/2022	27/05/2022
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	27/05/2022	27/05/2022
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	27/05/2022	27/05/2022
					CT449	Generators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	27/05/2022	27/05/2022
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	27/05/2022	27/05/2022
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	27/05/2022	27/05/2022
					CT981	Single use	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	27/05/2022	27/05/2022
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	27/05/2022	27/05/2022
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	27/05/2022	27/05/2022
					CT983	Surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	27/05/2022	27/05/2022
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	27/05/2022	27/05/2022
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	27/05/2022	27/05/2022
6041	2022	All other reportable incidents	47265	Dual-chamber implantable pacemaker, rate-responsive	CT141	Electrodes/Leads and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070101	Failure to Capture		
									A0712	Pacing Problem		
							F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A070101	Failure to Capture		
									A0712	Pacing Problem		
					CT145	Cardiovascular devices	E2403	No Clinical Signs, Symptoms or Conditions	A070101	Failure to Capture		
									A0712	Pacing Problem		
							F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A070101	Failure to Capture		
									A0712	Pacing Problem		
					CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070101	Failure to Capture		
									A0712	Pacing Problem		
							F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A070101	Failure to Capture		
									A0712	Pacing Problem		
					CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A070101	Failure to Capture		
									A0712	Pacing Problem		
							F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A070101	Failure to Capture		
									A0712	Pacing Problem		
					CT233	Surgical	E2403	No Clinical Signs, Symptoms or Conditions	A070101	Failure to Capture		
									A0712	Pacing Problem		
							F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A070101	Failure to Capture		
									A0712	Pacing Problem		
					CT2406	Implantable	E2403	No Clinical Signs, Symptoms or Conditions	A070101	Failure to Capture		
									A0712	Pacing Problem		
							F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A070101	Failure to Capture		
									A0712	Pacing Problem		
					CT2407	Active implantable	E2403	No Clinical Signs, Symptoms or Conditions	A070101	Failure to Capture		
									A0712	Pacing Problem		
							F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A070101	Failure to Capture		
									A0712	Pacing Problem		
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A070101	Failure to Capture		
									A0712	Pacing Problem		
							F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A070101	Failure to Capture		
									A0712	Pacing Problem		
					CT2727	Active implant devices/ancillaries	E2403	No Clinical Signs, Symptoms or Conditions	A070101	Failure to Capture		
									A0712	Pacing Problem		
							F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A070101	Failure to Capture		
									A0712	Pacing Problem		
					CT2750	Implantable cardiac electrical stimulators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070101	Failure to Capture		
									A0712	Pacing Problem		
							F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A070101	Failure to Capture		
									A0712	Pacing Problem		
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A070101	Failure to Capture		
									A0712	Pacing Problem		
							F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A070101	Failure to Capture		
									A0712	Pacing Problem		
					CT2985	Active	E2403	No Clinical Signs, Symptoms or Conditions	A070101	Failure to Capture		
									A0712	Pacing Problem		
							F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A070101	Failure to Capture		
									A0712	Pacing Problem		
					CT3001	Implantable metal	E2403	No Clinical Signs, Symptoms or Conditions	A070101	Failure to Capture		
									A0712	Pacing Problem		
							F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A070101	Failure to Capture		
									A0712	Pacing Problem		
					CT3002	Implantable synthetic polymer	E2403	No Clinical Signs, Symptoms or Conditions	A070101	Failure to Capture		
									A0712	Pacing Problem		
							F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A070101	Failure to Capture		
									A0712	Pacing Problem		
					CT3006	Implantable inorganic material	E2403	No Clinical Signs, Symptoms or Conditions	A070101	Failure to Capture		
									A0712	Pacing Problem		
							F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A070101	Failure to Capture		
									A0712	Pacing Problem		
					CT314	Electrically powered	E2403	No Clinical Signs, Symptoms or Conditions	A070101	Failure to Capture		
									A0712	Pacing Problem		
							F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A070101	Failure to Capture		
									A0712	Pacing Problem		
					CT321	Long-term surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A070101	Failure to Capture		
									A0712	Pacing Problem		
							F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A070101	Failure to Capture		
									A0712	Pacing Problem		

					CT449	Generators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070101	Failure to Capture		
									A0712	Pacing Problem		
							F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A070101	Failure to Capture		
									A0712	Pacing Problem		
					CT981	Single use	E2403	No Clinical Signs, Symptoms or Conditions	A070101	Failure to Capture		
									A0712	Pacing Problem		
							F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A070101	Failure to Capture		
									A0712	Pacing Problem		
					CT983	Surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A070101	Failure to Capture		
									A0712	Pacing Problem		
							F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A070101	Failure to Capture		
									A0712	Pacing Problem		
6042	2022	All other reportable incidents	35852	Single-chamber implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E0601	Arrhythmia	A070910	Undersensing		
									A071202	Inaccurate Synchronization		
							E2403	No Clinical Signs, Symptoms or Conditions	A070910	Undersensing		
									A071202	Inaccurate Synchronization		
							F08	Hospitalization or Prolonged Hospitalization	A070910	Undersensing		
									A071202	Inaccurate Synchronization		
							F15	Recognized Device or Procedural Complication	A070910	Undersensing		
									A071202	Inaccurate Synchronization		
6043	2022	All other reportable incidents	47265	Dual-chamber implantable pacemaker, rate-responsive	CT141	Electrodes/Leads and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070910	Undersensing		
									A0712	Pacing Problem		
									A1301	Failure to Read Input Signal		
							F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A070910	Undersensing		
									A0712	Pacing Problem		
									A1301	Failure to Read Input Signal		
							F15	Recognized Device or Procedural Complication	A070910	Undersensing		
									A0712	Pacing Problem		
									A1301	Failure to Read Input Signal		
					CT145	Cardiovascular devices	E2403	No Clinical Signs, Symptoms or Conditions	A070910	Undersensing		
									A0712	Pacing Problem		
									A1301	Failure to Read Input Signal		
							F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A070910	Undersensing		
									A0712	Pacing Problem		
									A1301	Failure to Read Input Signal		
							F15	Recognized Device or Procedural Complication	A070910	Undersensing		
									A0712	Pacing Problem		
									A1301	Failure to Read Input Signal		
					CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070910	Undersensing		
									A0712	Pacing Problem		
									A1301	Failure to Read Input Signal		
							F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A070910	Undersensing		
									A0712	Pacing Problem		
									A1301	Failure to Read Input Signal		
							F15	Recognized Device or Procedural Complication	A070910	Undersensing		
									A0712	Pacing Problem		
									A1301	Failure to Read Input Signal		
					CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A070910	Undersensing		
									A0712	Pacing Problem		
									A1301	Failure to Read Input Signal		
							F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A070910	Undersensing		
									A0712	Pacing Problem		
									A1301	Failure to Read Input Signal		
							F15	Recognized Device or Procedural Complication	A070910	Undersensing		
									A0712	Pacing Problem		
									A1301	Failure to Read Input Signal		
					CT233	Surgical	E2403	No Clinical Signs, Symptoms or Conditions	A070910	Undersensing		
									A0712	Pacing Problem		
									A1301	Failure to Read Input Signal		
							F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A070910	Undersensing		
									A0712	Pacing Problem		
									A1301	Failure to Read Input Signal		
							F15	Recognized Device or Procedural Complication	A070910	Undersensing		
									A0712	Pacing Problem		
									A1301	Failure to Read Input Signal		
					CT2406	Implantable	E2403	No Clinical Signs, Symptoms or Conditions	A070910	Undersensing		
									A0712	Pacing Problem		
									A1301	Failure to Read Input Signal		
							F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A070910	Undersensing		
									A0712	Pacing Problem		
									A1301	Failure to Read Input Signal		
							F15	Recognized Device or Procedural Complication	A070910	Undersensing		
									A0712	Pacing Problem		
									A1301	Failure to Read Input Signal		
					CT2407	Active implantable	E2403	No Clinical Signs, Symptoms or Conditions	A070910	Undersensing		
									A0712	Pacing Problem		
									A1301	Failure to Read Input Signal		
							F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A070910	Undersensing		
									A0712	Pacing Problem		
									A1301	Failure to Read Input Signal		
							F15	Recognized Device or Procedural Complication	A070910	Undersensing		
									A0712	Pacing Problem		
									A1301	Failure to Read Input Signal		
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A070910	Undersensing		
									A0712	Pacing Problem		
									A1301	Failure to Read Input Signal		
							F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A070910	Undersensing		
									A0712	Pacing Problem		

									A1301	Failure to Read Input Signal		
						F15	Recognized Device or Procedural Complication	A070910	Undersensing			
								A0712	Pacing Problem			
								A1301	Failure to Read Input Signal			
					CT2727	Active implant devices/ancillaries	E2403	No Clinical Signs, Symptoms or Conditions	A070910	Undersensing		
									A0712	Pacing Problem		
									A1301	Failure to Read Input Signal		
							F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A070910	Undersensing		
									A0712	Pacing Problem		
									A1301	Failure to Read Input Signal		
							F15	Recognized Device or Procedural Complication	A070910	Undersensing		
									A0712	Pacing Problem		
									A1301	Failure to Read Input Signal		
					CT2750	Implantable cardiac electrical stimulators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070910	Undersensing		
									A0712	Pacing Problem		
									A1301	Failure to Read Input Signal		
							F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A070910	Undersensing		
									A0712	Pacing Problem		
									A1301	Failure to Read Input Signal		
							F15	Recognized Device or Procedural Complication	A070910	Undersensing		
									A0712	Pacing Problem		
									A1301	Failure to Read Input Signal		
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A070910	Undersensing		
									A0712	Pacing Problem		
									A1301	Failure to Read Input Signal		
							F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A070910	Undersensing		
									A0712	Pacing Problem		
									A1301	Failure to Read Input Signal		
							F15	Recognized Device or Procedural Complication	A070910	Undersensing		
									A0712	Pacing Problem		
									A1301	Failure to Read Input Signal		
					CT2985	Active	E2403	No Clinical Signs, Symptoms or Conditions	A070910	Undersensing		
									A0712	Pacing Problem		
									A1301	Failure to Read Input Signal		
							F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A070910	Undersensing		
									A0712	Pacing Problem		
									A1301	Failure to Read Input Signal		
							F15	Recognized Device or Procedural Complication	A070910	Undersensing		
									A0712	Pacing Problem		
									A1301	Failure to Read Input Signal		
					CT3001	Implantable metal	E2403	No Clinical Signs, Symptoms or Conditions	A070910	Undersensing		
									A0712	Pacing Problem		
									A1301	Failure to Read Input Signal		
							F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A070910	Undersensing		
									A0712	Pacing Problem		
									A1301	Failure to Read Input Signal		
							F15	Recognized Device or Procedural Complication	A070910	Undersensing		
									A0712	Pacing Problem		
									A1301	Failure to Read Input Signal		
					CT3002	Implantable synthetic polymer	E2403	No Clinical Signs, Symptoms or Conditions	A070910	Undersensing		
									A0712	Pacing Problem		
									A1301	Failure to Read Input Signal		
							F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A070910	Undersensing		
									A0712	Pacing Problem		
									A1301	Failure to Read Input Signal		
							F15	Recognized Device or Procedural Complication	A070910	Undersensing		
									A0712	Pacing Problem		
									A1301	Failure to Read Input Signal		
					CT3006	Implantable inorganic material	E2403	No Clinical Signs, Symptoms or Conditions	A070910	Undersensing		
									A0712	Pacing Problem		
									A1301	Failure to Read Input Signal		
							F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A070910	Undersensing		
									A0712	Pacing Problem		
									A1301	Failure to Read Input Signal		
							F15	Recognized Device or Procedural Complication	A070910	Undersensing		
									A0712	Pacing Problem		
									A1301	Failure to Read Input Signal		
					CT314	Electrically powered	E2403	No Clinical Signs, Symptoms or Conditions	A070910	Undersensing		
									A0712	Pacing Problem		
									A1301	Failure to Read Input Signal		
							F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A070910	Undersensing		
									A0712	Pacing Problem		
									A1301	Failure to Read Input Signal		
							F15	Recognized Device or Procedural Complication	A070910	Undersensing		
									A0712	Pacing Problem		
									A1301	Failure to Read Input Signal		
					CT321	Long-term surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A070910	Undersensing		
									A0712	Pacing Problem		
									A1301	Failure to Read Input Signal		
							F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A070910	Undersensing		
									A0712	Pacing Problem		
									A1301	Failure to Read Input Signal		
							F15	Recognized Device or Procedural Complication	A070910	Undersensing		
									A0712	Pacing Problem		
									A1301	Failure to Read Input Signal		
					CT449	Generators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070910	Undersensing		
									A0712	Pacing Problem		
									A1301	Failure to Read Input Signal		

						F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A070910	Undersensing			
								A0712	Pacing Problem			
								A1301	Failure to Read Input Signal			
						F15	Recognized Device or Procedural Complication	A070910	Undersensing			
								A0712	Pacing Problem			
								A1301	Failure to Read Input Signal			
					CT981	Single use	E2403	No Clinical Signs, Symptoms or Conditions	A070910	Undersensing		
									A0712	Pacing Problem		
									A1301	Failure to Read Input Signal		
							F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A070910	Undersensing		
									A0712	Pacing Problem		
									A1301	Failure to Read Input Signal		
							F15	Recognized Device or Procedural Complication	A070910	Undersensing		
									A0712	Pacing Problem		
									A1301	Failure to Read Input Signal		
					CT983	Surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A070910	Undersensing		
									A0712	Pacing Problem		
									A1301	Failure to Read Input Signal		
							F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A070910	Undersensing		
									A0712	Pacing Problem		
									A1301	Failure to Read Input Signal		
							F15	Recognized Device or Procedural Complication	A070910	Undersensing		
									A0712	Pacing Problem		
									A1301	Failure to Read Input Signal		
6044	2022	All other reportable incidents	47270	Cardiac resynchronization therapy implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E2104	Electric Shock	A070909	Oversensing		
									A071302	Inappropriate/Inadequate Shock/Stimulation		
							E2403	No Clinical Signs, Symptoms or Conditions	A070909	Oversensing		
									A071302	Inappropriate/Inadequate Shock/Stimulation		
							F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A070909	Oversensing		
									A071302	Inappropriate/Inadequate Shock/Stimulation		
6045	2021	All other reportable incidents	35223	Endocardial/interventricular septal pacing lead	CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070101	Failure to Capture		
									A072201	High Impedance		
							F26	No Health Consequences or Impact	A070101	Failure to Capture		
									A072201	High Impedance		
6046	2021	All other reportable incidents	17882	Physiologic-monitoring defibrillation system	CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem		
							F27	Problem Identified During Non-Clinical Procedure	A0705	Battery Problem		
6047	2021	All other reportable incidents	37265	Dual-chamber implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem	28/07/2021	28/07/2021
									A070504	Premature Discharge of Battery	28/07/2021	28/07/2021
							F08	Hospitalization or Prolonged Hospitalization	A0705	Battery Problem	28/07/2021	28/07/2021
									A070504	Premature Discharge of Battery	28/07/2021	28/07/2021
							F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A0705	Battery Problem	28/07/2021	28/07/2021
									A070504	Premature Discharge of Battery	28/07/2021	28/07/2021
							F1905	Device Revision or Replacement	A0705	Battery Problem	28/07/2021	28/07/2021
									A070504	Premature Discharge of Battery	28/07/2021	28/07/2021
6048	2021	All other reportable incidents	47265	Dual-chamber implantable pacemaker, rate-responsive	CT141	Electrodes/Leads and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F15	Recognized Device or Procedural Complication	A070504	Premature Discharge of Battery		
					CT145	Cardiovascular devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F15	Recognized Device or Procedural Complication	A070504	Premature Discharge of Battery		
					CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F15	Recognized Device or Procedural Complication	A070504	Premature Discharge of Battery		
					CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F15	Recognized Device or Procedural Complication	A070504	Premature Discharge of Battery		
					CT233	Surgical	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F15	Recognized Device or Procedural Complication	A070504	Premature Discharge of Battery		
					CT2406	Implantable	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F15	Recognized Device or Procedural Complication	A070504	Premature Discharge of Battery		
					CT2407	Active implantable	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F15	Recognized Device or Procedural Complication	A070504	Premature Discharge of Battery		
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F15	Recognized Device or Procedural Complication	A070504	Premature Discharge of Battery		
					CT2727	Active implant devices/ancillaries	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F15	Recognized Device or Procedural Complication	A070504	Premature Discharge of Battery		
					CT2750	Implantable cardiac electrical stimulators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F15	Recognized Device or Procedural Complication	A070504	Premature Discharge of Battery		
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F15	Recognized Device or Procedural Complication	A070504	Premature Discharge of Battery		
					CT2985	Active	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F15	Recognized Device or Procedural Complication	A070504	Premature Discharge of Battery		
					CT3001	Implantable metal	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F15	Recognized Device or Procedural Complication	A070504	Premature Discharge of Battery		
					CT3002	Implantable synthetic polymer	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F15	Recognized Device or Procedural Complication	A070504	Premature Discharge of Battery		
					CT3006	Implantable inorganic material	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F15	Recognized Device or Procedural Complication	A070504	Premature Discharge of Battery		
					CT314	Electrically powered	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F15	Recognized Device or Procedural Complication	A070504	Premature Discharge of Battery		
					CT321	Long-term surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F15	Recognized Device or Procedural Complication	A070504	Premature Discharge of Battery		
					CT449	Generators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F15	Recognized Device or Procedural Complication	A070504	Premature Discharge of Battery		
					CT981	Single use	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F15	Recognized Device or Procedural Complication	A070504	Premature Discharge of Battery		
					CT983	Surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F15	Recognized Device or Procedural Complication	A070504	Premature Discharge of Battery		
6049	2021	All other reportable incidents	47267	Single-chamber implantable pacemaker, rate-responsive	CT141	Electrodes/Leads and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	29/07/2021	29/07/2021
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	29/07/2021	29/07/2021
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	29/07/2021	29/07/2021
					CT145	Cardiovascular devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	29/07/2021	29/07/2021

							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	29/07/2021	29/07/2021
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	29/07/2021	29/07/2021
					CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	29/07/2021	29/07/2021
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	29/07/2021	29/07/2021
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	29/07/2021	29/07/2021
					CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	29/07/2021	29/07/2021
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	29/07/2021	29/07/2021
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	29/07/2021	29/07/2021
					CT233	Surgical	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	29/07/2021	29/07/2021
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	29/07/2021	29/07/2021
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	29/07/2021	29/07/2021
					CT2406	Implantable	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	29/07/2021	29/07/2021
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	29/07/2021	29/07/2021
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	29/07/2021	29/07/2021
					CT2407	Active implantable	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	29/07/2021	29/07/2021
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	29/07/2021	29/07/2021
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	29/07/2021	29/07/2021
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	29/07/2021	29/07/2021
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	29/07/2021	29/07/2021
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	29/07/2021	29/07/2021
					CT2727	Active implant devices/ancillaries	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	29/07/2021	29/07/2021
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	29/07/2021	29/07/2021
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	29/07/2021	29/07/2021
					CT2750	Implantable cardiac electrical stimulators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	29/07/2021	29/07/2021
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	29/07/2021	29/07/2021
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	29/07/2021	29/07/2021
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	29/07/2021	29/07/2021
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	29/07/2021	29/07/2021
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	29/07/2021	29/07/2021
					CT2985	Active	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	29/07/2021	29/07/2021
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	29/07/2021	29/07/2021
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	29/07/2021	29/07/2021
					CT3001	Implantable metal	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	29/07/2021	29/07/2021
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	29/07/2021	29/07/2021
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	29/07/2021	29/07/2021
					CT3002	Implantable synthetic polymer	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	29/07/2021	29/07/2021
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	29/07/2021	29/07/2021
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	29/07/2021	29/07/2021
					CT3006	Implantable inorganic material	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	29/07/2021	29/07/2021
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	29/07/2021	29/07/2021
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	29/07/2021	29/07/2021
					CT314	Electrically powered	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	29/07/2021	29/07/2021
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	29/07/2021	29/07/2021
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	29/07/2021	29/07/2021
					CT321	Long-term surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	29/07/2021	29/07/2021
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	29/07/2021	29/07/2021
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	29/07/2021	29/07/2021
					CT449	Generators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	29/07/2021	29/07/2021
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	29/07/2021	29/07/2021
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	29/07/2021	29/07/2021
					CT981	Single use	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	29/07/2021	29/07/2021
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	29/07/2021	29/07/2021
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	29/07/2021	29/07/2021
					CT983	Surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	29/07/2021	29/07/2021
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	29/07/2021	29/07/2021
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	29/07/2021	29/07/2021
6050	2021	All other reportable incidents	35852	Single-chamber implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem	27/09/2021	27/09/2021
									A070504	Premature Discharge of Battery	27/09/2021	27/09/2021
							F08	Hospitalization or Prolonged Hospitalization	A0705	Battery Problem	27/09/2021	27/09/2021
									A070504	Premature Discharge of Battery	27/09/2021	27/09/2021
							F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A0705	Battery Problem	27/09/2021	27/09/2021
									A070504	Premature Discharge of Battery	27/09/2021	27/09/2021
							F1905	Device Revision or Replacement	A0705	Battery Problem	27/09/2021	27/09/2021
									A070504	Premature Discharge of Battery	27/09/2021	27/09/2021
6051	2021	All other reportable incidents	35852	Single-chamber implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A051201	Device Dislodged or Dislocated		
									A070909	Oversensing		
									A090801	Signal Artifact/Noise		
									A1301	Failure to Read Input Signal		
							F12	Serious Injury/ Illness/ Impairment	A051201	Device Dislodged or Dislocated		
									A070909	Oversensing		
									A090801	Signal Artifact/Noise		
									A1301	Failure to Read Input Signal		
							F13	Misdiagnosis/ Misclassification	A051201	Device Dislodged or Dislocated		
									A070909	Oversensing		
									A090801	Signal Artifact/Noise		
									A1301	Failure to Read Input Signal		
							F19	Surgical Intervention	A051201	Device Dislodged or Dislocated		
									A070909	Oversensing		
									A090801	Signal Artifact/Noise		
									A1301	Failure to Read Input Signal		
6052	2021	All other reportable incidents	47267	Single-chamber implantable pacemaker, rate-responsive	CT141	Electrodes/Leads and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
					CT145	Cardiovascular devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
					CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
					CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
					CT233	Surgical	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
					CT2406	Implantable	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
					CT2407	Active implantable	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		

					CT2727	Active implant devices/ancillaries	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
					CT2750	Implantable cardiac electrical stimulators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
					CT2985	Active	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
					CT3001	Implantable metal	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
					CT3002	Implantable synthetic polymer	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
					CT3006	Implantable inorganic material	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
					CT314	Electrically powered	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
					CT321	Long-term surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
					CT449	Generators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
					CT981	Single use	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
					CT983	Surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
6053	2022	All other reportable incidents	35852	Single-chamber implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A070909	Oversensing		
							F23	Unexpected Medical Intervention	A070909	Oversensing		
6054	2021	All other reportable incidents	35852	Single-chamber implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem	27/08/2021	27/08/2021
									A070504	Premature Discharge of Battery	27/08/2021	27/08/2021
							F08	Hospitalization or Prolonged Hospitalization	A0705	Battery Problem	27/08/2021	27/08/2021
									A070504	Premature Discharge of Battery	27/08/2021	27/08/2021
							F1905	Device Revision or Replacement	A0705	Battery Problem	27/08/2021	27/08/2021
									A070504	Premature Discharge of Battery	27/08/2021	27/08/2021
6055	2021	All other reportable incidents	47263	Cardiac resynchronization therapy implantable pacemaker	CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A0908	Incorrect, Inadequate or Imprecise Result or Readings	09/07/2021	09/07/2021
							F08	Hospitalization or Prolonged Hospitalization	A0908	Incorrect, Inadequate or Imprecise Result or Readings	09/07/2021	09/07/2021
							F1905	Device Revision or Replacement	A0908	Incorrect, Inadequate or Imprecise Result or Readings	09/07/2021	09/07/2021
					CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A0908	Incorrect, Inadequate or Imprecise Result or Readings	09/07/2021	09/07/2021
							F08	Hospitalization or Prolonged Hospitalization	A0908	Incorrect, Inadequate or Imprecise Result or Readings	09/07/2021	09/07/2021
							F1905	Device Revision or Replacement	A0908	Incorrect, Inadequate or Imprecise Result or Readings	09/07/2021	09/07/2021
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A0908	Incorrect, Inadequate or Imprecise Result or Readings	09/07/2021	09/07/2021
							F08	Hospitalization or Prolonged Hospitalization	A0908	Incorrect, Inadequate or Imprecise Result or Readings	09/07/2021	09/07/2021
							F1905	Device Revision or Replacement	A0908	Incorrect, Inadequate or Imprecise Result or Readings	09/07/2021	09/07/2021
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A0908	Incorrect, Inadequate or Imprecise Result or Readings	09/07/2021	09/07/2021
							F08	Hospitalization or Prolonged Hospitalization	A0908	Incorrect, Inadequate or Imprecise Result or Readings	09/07/2021	09/07/2021
							F1905	Device Revision or Replacement	A0908	Incorrect, Inadequate or Imprecise Result or Readings	09/07/2021	09/07/2021
6056	2021	All other reportable incidents	35852	Single-chamber implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A0508	Noise, Audible	18/10/2021	18/10/2021
									A070504	Premature Discharge of Battery	18/10/2021	18/10/2021
							F08	Hospitalization or Prolonged Hospitalization	A0508	Noise, Audible	18/10/2021	18/10/2021
									A070504	Premature Discharge of Battery	18/10/2021	18/10/2021
							F1905	Device Revision or Replacement	A0508	Noise, Audible	18/10/2021	18/10/2021
									A070504	Premature Discharge of Battery	18/10/2021	18/10/2021
6057	2021	All other reportable incidents	37265	Dual-chamber implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A04	Material Integrity Problem	02/07/2021	02/07/2021
							F1905	Device Revision or Replacement	A04	Material Integrity Problem	02/07/2021	02/07/2021
							F26	No Health Consequences or Impact	A04	Material Integrity Problem	02/07/2021	02/07/2021
6058	2022	All other reportable incidents	35852	Single-chamber implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
6059	2022	All other reportable incidents	36241	Pacemaker/defibrillator lead	CT1684	Cardiac pacemakers and associated devices	E0601	Arrhythmia	A24	No Apparent Device Problem	03/02/2022	03/02/2022
							E0604	Cardiac Perforation	A24	No Apparent Device Problem	03/02/2022	03/02/2022
							E2104	Electric Shock	A24	No Apparent Device Problem	03/02/2022	03/02/2022
							E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	03/02/2022	03/02/2022
							F1908	Prolonged Surgery	A24	No Apparent Device Problem	03/02/2022	03/02/2022
							F2306	Resuscitation	A24	No Apparent Device Problem	03/02/2022	03/02/2022
6060	2022	All other reportable incidents	35223	Endocardial/interventricular septal pacing lead	CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070909	Oversensing		
							F19	Surgical Intervention	A070909	Oversensing		
6061	2022	All other reportable incidents	47267	Single-chamber implantable pacemaker, rate-responsive	CT141	Electrodes/Leads and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem		
							F26	No Health Consequences or Impact	A0705	Battery Problem		
					CT145	Cardiovascular devices	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem		
							F26	No Health Consequences or Impact	A0705	Battery Problem		
					CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem		
							F26	No Health Consequences or Impact	A0705	Battery Problem		
					CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem		
							F26	No Health Consequences or Impact	A0705	Battery Problem		
					CT233	Surgical	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem		
							F26	No Health Consequences or Impact	A0705	Battery Problem		
					CT2406	Implantable	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem		
							F26	No Health Consequences or Impact	A0705	Battery Problem		
					CT2407	Active implantable	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem		
							F26	No Health Consequences or Impact	A0705	Battery Problem		
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem		
							F26	No Health Consequences or Impact	A0705	Battery Problem		
					CT2727	Active implant devices/ancillaries	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem		
							F26	No Health Consequences or Impact	A0705	Battery Problem		
					CT2750	Implantable cardiac electrical stimulators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem		
							F26	No Health Consequences or Impact	A0705	Battery Problem		
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem		
							F26	No Health Consequences or Impact	A0705	Battery Problem		
					CT2985	Active	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem		
							F26	No Health Consequences or Impact	A0705	Battery Problem		
					CT3001	Implantable metal	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem		

						F26	No Health Consequences or Impact	A0705	Battery Problem		
				CT3002	Implantable synthetic polymer	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem		
						F26	No Health Consequences or Impact	A0705	Battery Problem		
				CT3006	Implantable inorganic material	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem		
						F26	No Health Consequences or Impact	A0705	Battery Problem		
				CT314	Electrically powered	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem		
						F26	No Health Consequences or Impact	A0705	Battery Problem		
				CT321	Long-term surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem		
						F26	No Health Consequences or Impact	A0705	Battery Problem		
				CT449	Generators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem		
						F26	No Health Consequences or Impact	A0705	Battery Problem		
				CT981	Single use	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem		
						F26	No Health Consequences or Impact	A0705	Battery Problem		
				CT983	Surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem		
						F26	No Health Consequences or Impact	A0705	Battery Problem		
6062	2021	All other reportable incidents	37265	Dual-chamber implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A070603	Failure to Charge	12/07/2021	12/07/2021
							F1901	Additional Surgery	A070603	Failure to Charge	12/07/2021	12/07/2021
6063	2021	All other reportable incidents	47265	Dual-chamber implantable pacemaker, rate-responsive	CT141	Electrodes/Leads and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	12/07/2021	12/07/2021
							F1901	Additional Surgery	A070504	Premature Discharge of Battery	12/07/2021	12/07/2021
					CT145	Cardiovascular devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	12/07/2021	12/07/2021
							F1901	Additional Surgery	A070504	Premature Discharge of Battery	12/07/2021	12/07/2021
					CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	12/07/2021	12/07/2021
							F1901	Additional Surgery	A070504	Premature Discharge of Battery	12/07/2021	12/07/2021
					CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	12/07/2021	12/07/2021
							F1901	Additional Surgery	A070504	Premature Discharge of Battery	12/07/2021	12/07/2021
					CT233	Surgical	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	12/07/2021	12/07/2021
							F1901	Additional Surgery	A070504	Premature Discharge of Battery	12/07/2021	12/07/2021
					CT2406	Implantable	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	12/07/2021	12/07/2021
							F1901	Additional Surgery	A070504	Premature Discharge of Battery	12/07/2021	12/07/2021
					CT2407	Active implantable	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	12/07/2021	12/07/2021
							F1901	Additional Surgery	A070504	Premature Discharge of Battery	12/07/2021	12/07/2021
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	12/07/2021	12/07/2021
							F1901	Additional Surgery	A070504	Premature Discharge of Battery	12/07/2021	12/07/2021
					CT2727	Active implant devices/ancillaries	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	12/07/2021	12/07/2021
							F1901	Additional Surgery	A070504	Premature Discharge of Battery	12/07/2021	12/07/2021
					CT2750	Implantable cardiac electrical stimulators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	12/07/2021	12/07/2021
							F1901	Additional Surgery	A070504	Premature Discharge of Battery	12/07/2021	12/07/2021
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	12/07/2021	12/07/2021
							F1901	Additional Surgery	A070504	Premature Discharge of Battery	12/07/2021	12/07/2021
					CT2985	Active	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	12/07/2021	12/07/2021
							F1901	Additional Surgery	A070504	Premature Discharge of Battery	12/07/2021	12/07/2021
					CT3001	Implantable metal	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	12/07/2021	12/07/2021
							F1901	Additional Surgery	A070504	Premature Discharge of Battery	12/07/2021	12/07/2021
					CT3002	Implantable synthetic polymer	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	12/07/2021	12/07/2021
							F1901	Additional Surgery	A070504	Premature Discharge of Battery	12/07/2021	12/07/2021
					CT3006	Implantable inorganic material	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	12/07/2021	12/07/2021
							F1901	Additional Surgery	A070504	Premature Discharge of Battery	12/07/2021	12/07/2021
					CT314	Electrically powered	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	12/07/2021	12/07/2021
							F1901	Additional Surgery	A070504	Premature Discharge of Battery	12/07/2021	12/07/2021
					CT321	Long-term surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	12/07/2021	12/07/2021
							F1901	Additional Surgery	A070504	Premature Discharge of Battery	12/07/2021	12/07/2021
					CT449	Generators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	12/07/2021	12/07/2021
							F1901	Additional Surgery	A070504	Premature Discharge of Battery	12/07/2021	12/07/2021
					CT981	Single use	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	12/07/2021	12/07/2021
							F1901	Additional Surgery	A070504	Premature Discharge of Battery	12/07/2021	12/07/2021
					CT983	Surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	12/07/2021	12/07/2021
							F1901	Additional Surgery	A070504	Premature Discharge of Battery	12/07/2021	12/07/2021
6064	2021	Death	17882	Physiologic-monitoring defibrillation system	CT1684	Cardiac pacemakers and associated devices	E0602	Cardiac Arrest	A090811	Unable to Obtain Readings		
							F02	Death	A090811	Unable to Obtain Readings		
6065	2022	All other reportable incidents	47267	Single-chamber implantable pacemaker, rate-responsive	CT141	Electrodes/Leads and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	11/05/2022	11/05/2022
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	11/05/2022	11/05/2022
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	11/05/2022	11/05/2022
					CT145	Cardiovascular devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	11/05/2022	11/05/2022
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	11/05/2022	11/05/2022
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	11/05/2022	11/05/2022
					CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	11/05/2022	11/05/2022
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	11/05/2022	11/05/2022
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	11/05/2022	11/05/2022
					CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	11/05/2022	11/05/2022
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	11/05/2022	11/05/2022
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	11/05/2022	11/05/2022
					CT233	Surgical	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	11/05/2022	11/05/2022
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	11/05/2022	11/05/2022
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	11/05/2022	11/05/2022
					CT2406	Implantable	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	11/05/2022	11/05/2022
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	11/05/2022	11/05/2022
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	11/05/2022	11/05/2022
					CT2407	Active implantable	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	11/05/2022	11/05/2022
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	11/05/2022	11/05/2022
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	11/05/2022	11/05/2022
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	11/05/2022	11/05/2022
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	11/05/2022	11/05/2022
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	11/05/2022	11/05/2022
					CT2727	Active implant devices/ancillaries	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	11/05/2022	11/05/2022
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	11/05/2022	11/05/2022
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	11/05/2022	11/05/2022
					CT2750	Implantable cardiac electrical stimulators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	11/05/2022	11/05/2022
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	11/05/2022	11/05/2022

							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	11/05/2022	11/05/2022
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	11/05/2022	11/05/2022
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	11/05/2022	11/05/2022
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	11/05/2022	11/05/2022
					CT2985	Active	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	11/05/2022	11/05/2022
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	11/05/2022	11/05/2022
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	11/05/2022	11/05/2022
					CT3001	Implantable metal	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	11/05/2022	11/05/2022
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	11/05/2022	11/05/2022
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	11/05/2022	11/05/2022
					CT3002	Implantable synthetic polymer	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	11/05/2022	11/05/2022
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	11/05/2022	11/05/2022
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	11/05/2022	11/05/2022
					CT3006	Implantable inorganic material	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	11/05/2022	11/05/2022
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	11/05/2022	11/05/2022
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	11/05/2022	11/05/2022
					CT314	Electrically powered	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	11/05/2022	11/05/2022
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	11/05/2022	11/05/2022
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	11/05/2022	11/05/2022
					CT321	Long-term surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	11/05/2022	11/05/2022
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	11/05/2022	11/05/2022
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	11/05/2022	11/05/2022
					CT449	Generators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	11/05/2022	11/05/2022
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	11/05/2022	11/05/2022
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	11/05/2022	11/05/2022
					CT981	Single use	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	11/05/2022	11/05/2022
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	11/05/2022	11/05/2022
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	11/05/2022	11/05/2022
					CT983	Surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	11/05/2022	11/05/2022
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	11/05/2022	11/05/2022
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	11/05/2022	11/05/2022
6066	2022	All other reportable incidents	47265	Dual-chamber implantable pacemaker, rate-responsive	CT141	Electrodes/Leads and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem	11/07/2022	11/07/2022
									A070504	Premature Discharge of Battery	11/07/2022	11/07/2022
							F08	Hospitalization or Prolonged Hospitalization	A0705	Battery Problem	11/07/2022	11/07/2022
									A070504	Premature Discharge of Battery	11/07/2022	11/07/2022
							F1903	Device Explantation	A0705	Battery Problem	11/07/2022	11/07/2022
									A070504	Premature Discharge of Battery	11/07/2022	11/07/2022
					CT145	Cardiovascular devices	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem	11/07/2022	11/07/2022
									A070504	Premature Discharge of Battery	11/07/2022	11/07/2022
							F08	Hospitalization or Prolonged Hospitalization	A0705	Battery Problem	11/07/2022	11/07/2022
									A070504	Premature Discharge of Battery	11/07/2022	11/07/2022
							F1903	Device Explantation	A0705	Battery Problem	11/07/2022	11/07/2022
									A070504	Premature Discharge of Battery	11/07/2022	11/07/2022
					CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem	11/07/2022	11/07/2022
									A070504	Premature Discharge of Battery	11/07/2022	11/07/2022
							F08	Hospitalization or Prolonged Hospitalization	A0705	Battery Problem	11/07/2022	11/07/2022
									A070504	Premature Discharge of Battery	11/07/2022	11/07/2022
							F1903	Device Explantation	A0705	Battery Problem	11/07/2022	11/07/2022
									A070504	Premature Discharge of Battery	11/07/2022	11/07/2022
					CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem	11/07/2022	11/07/2022
									A070504	Premature Discharge of Battery	11/07/2022	11/07/2022
							F08	Hospitalization or Prolonged Hospitalization	A0705	Battery Problem	11/07/2022	11/07/2022
									A070504	Premature Discharge of Battery	11/07/2022	11/07/2022
							F1903	Device Explantation	A0705	Battery Problem	11/07/2022	11/07/2022
									A070504	Premature Discharge of Battery	11/07/2022	11/07/2022
					CT233	Surgical	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem	11/07/2022	11/07/2022
									A070504	Premature Discharge of Battery	11/07/2022	11/07/2022
							F08	Hospitalization or Prolonged Hospitalization	A0705	Battery Problem	11/07/2022	11/07/2022
									A070504	Premature Discharge of Battery	11/07/2022	11/07/2022
							F1903	Device Explantation	A0705	Battery Problem	11/07/2022	11/07/2022
									A070504	Premature Discharge of Battery	11/07/2022	11/07/2022
					CT2406	Implantable	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem	11/07/2022	11/07/2022
									A070504	Premature Discharge of Battery	11/07/2022	11/07/2022
							F08	Hospitalization or Prolonged Hospitalization	A0705	Battery Problem	11/07/2022	11/07/2022
									A070504	Premature Discharge of Battery	11/07/2022	11/07/2022
							F1903	Device Explantation	A0705	Battery Problem	11/07/2022	11/07/2022
									A070504	Premature Discharge of Battery	11/07/2022	11/07/2022
					CT2407	Active implantable	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem	11/07/2022	11/07/2022
									A070504	Premature Discharge of Battery	11/07/2022	11/07/2022
							F08	Hospitalization or Prolonged Hospitalization	A0705	Battery Problem	11/07/2022	11/07/2022
									A070504	Premature Discharge of Battery	11/07/2022	11/07/2022
							F1903	Device Explantation	A0705	Battery Problem	11/07/2022	11/07/2022
									A070504	Premature Discharge of Battery	11/07/2022	11/07/2022
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem	11/07/2022	11/07/2022
									A070504	Premature Discharge of Battery	11/07/2022	11/07/2022
							F08	Hospitalization or Prolonged Hospitalization	A0705	Battery Problem	11/07/2022	11/07/2022
									A070504	Premature Discharge of Battery	11/07/2022	11/07/2022
							F1903	Device Explantation	A0705	Battery Problem	11/07/2022	11/07/2022
									A070504	Premature Discharge of Battery	11/07/2022	11/07/2022
					CT2727	Active implant devices/ancillaries	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem	11/07/2022	11/07/2022
									A070504	Premature Discharge of Battery	11/07/2022	11/07/2022
							F08	Hospitalization or Prolonged Hospitalization	A0705	Battery Problem	11/07/2022	11/07/2022
									A070504	Premature Discharge of Battery	11/07/2022	11/07/2022
							F1903	Device Explantation	A0705	Battery Problem	11/07/2022	11/07/2022
									A070504	Premature Discharge of Battery	11/07/2022	11/07/2022
					CT2750	Implantable cardiac electrical stimulators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem	11/07/2022	11/07/2022
									A070504	Premature Discharge of Battery	11/07/2022	11/07/2022
							F08	Hospitalization or Prolonged Hospitalization	A0705	Battery Problem	11/07/2022	11/07/2022

								A070504	Premature Discharge of Battery	11/07/2022	11/07/2022
						F1903	Device Explantation	A0705	Battery Problem	11/07/2022	11/07/2022
								A070504	Premature Discharge of Battery	11/07/2022	11/07/2022
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem	11/07/2022	11/07/2022
									A070504	Premature Discharge of Battery	11/07/2022	11/07/2022
							F08	Hospitalization or Prolonged Hospitalization	A0705	Battery Problem	11/07/2022	11/07/2022
									A070504	Premature Discharge of Battery	11/07/2022	11/07/2022
							F1903	Device Explantation	A0705	Battery Problem	11/07/2022	11/07/2022
									A070504	Premature Discharge of Battery	11/07/2022	11/07/2022
					CT2985	Active	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem	11/07/2022	11/07/2022
									A070504	Premature Discharge of Battery	11/07/2022	11/07/2022
							F08	Hospitalization or Prolonged Hospitalization	A0705	Battery Problem	11/07/2022	11/07/2022
									A070504	Premature Discharge of Battery	11/07/2022	11/07/2022
							F1903	Device Explantation	A0705	Battery Problem	11/07/2022	11/07/2022
									A070504	Premature Discharge of Battery	11/07/2022	11/07/2022
					CT3001	Implantable metal	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem	11/07/2022	11/07/2022
									A070504	Premature Discharge of Battery	11/07/2022	11/07/2022
							F08	Hospitalization or Prolonged Hospitalization	A0705	Battery Problem	11/07/2022	11/07/2022
									A070504	Premature Discharge of Battery	11/07/2022	11/07/2022
							F1903	Device Explantation	A0705	Battery Problem	11/07/2022	11/07/2022
									A070504	Premature Discharge of Battery	11/07/2022	11/07/2022
					CT3002	Implantable synthetic polymer	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem	11/07/2022	11/07/2022
									A070504	Premature Discharge of Battery	11/07/2022	11/07/2022
							F08	Hospitalization or Prolonged Hospitalization	A0705	Battery Problem	11/07/2022	11/07/2022
									A070504	Premature Discharge of Battery	11/07/2022	11/07/2022
							F1903	Device Explantation	A0705	Battery Problem	11/07/2022	11/07/2022
									A070504	Premature Discharge of Battery	11/07/2022	11/07/2022
					CT3006	Implantable inorganic material	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem	11/07/2022	11/07/2022
									A070504	Premature Discharge of Battery	11/07/2022	11/07/2022
							F08	Hospitalization or Prolonged Hospitalization	A0705	Battery Problem	11/07/2022	11/07/2022
									A070504	Premature Discharge of Battery	11/07/2022	11/07/2022
							F1903	Device Explantation	A0705	Battery Problem	11/07/2022	11/07/2022
									A070504	Premature Discharge of Battery	11/07/2022	11/07/2022
					CT314	Electrically powered	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem	11/07/2022	11/07/2022
									A070504	Premature Discharge of Battery	11/07/2022	11/07/2022
							F08	Hospitalization or Prolonged Hospitalization	A0705	Battery Problem	11/07/2022	11/07/2022
									A070504	Premature Discharge of Battery	11/07/2022	11/07/2022
							F1903	Device Explantation	A0705	Battery Problem	11/07/2022	11/07/2022
									A070504	Premature Discharge of Battery	11/07/2022	11/07/2022
					CT321	Long-term surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem	11/07/2022	11/07/2022
									A070504	Premature Discharge of Battery	11/07/2022	11/07/2022
							F08	Hospitalization or Prolonged Hospitalization	A0705	Battery Problem	11/07/2022	11/07/2022
									A070504	Premature Discharge of Battery	11/07/2022	11/07/2022
							F1903	Device Explantation	A0705	Battery Problem	11/07/2022	11/07/2022
									A070504	Premature Discharge of Battery	11/07/2022	11/07/2022
					CT449	Generators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem	11/07/2022	11/07/2022
									A070504	Premature Discharge of Battery	11/07/2022	11/07/2022
							F08	Hospitalization or Prolonged Hospitalization	A0705	Battery Problem	11/07/2022	11/07/2022
									A070504	Premature Discharge of Battery	11/07/2022	11/07/2022
							F1903	Device Explantation	A0705	Battery Problem	11/07/2022	11/07/2022
									A070504	Premature Discharge of Battery	11/07/2022	11/07/2022
					CT981	Single use	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem	11/07/2022	11/07/2022
									A070504	Premature Discharge of Battery	11/07/2022	11/07/2022
							F08	Hospitalization or Prolonged Hospitalization	A0705	Battery Problem	11/07/2022	11/07/2022
									A070504	Premature Discharge of Battery	11/07/2022	11/07/2022
							F1903	Device Explantation	A0705	Battery Problem	11/07/2022	11/07/2022
									A070504	Premature Discharge of Battery	11/07/2022	11/07/2022
					CT983	Surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem	11/07/2022	11/07/2022
									A070504	Premature Discharge of Battery	11/07/2022	11/07/2022
							F08	Hospitalization or Prolonged Hospitalization	A0705	Battery Problem	11/07/2022	11/07/2022
									A070504	Premature Discharge of Battery	11/07/2022	11/07/2022
							F1903	Device Explantation	A0705	Battery Problem	11/07/2022	11/07/2022
									A070504	Premature Discharge of Battery	11/07/2022	11/07/2022
6067	2022	All other reportable incidents	36241	Pacemaker/defibrillator lead	CT1684	Cardiac pacemakers and associated devices	E060101	Asystole	A0203	Defective Device		
									A070101	Failure to Capture		
									A070909	Oversensing		
									A0712	Pacing Problem		
									A072201	High Impedance		
							E2403	No Clinical Signs, Symptoms or Conditions	A0203	Defective Device		
									A070101	Failure to Capture		
									A070909	Oversensing		
									A0712	Pacing Problem		
									A072201	High Impedance		
							F1001	Absence of Treatment	A0203	Defective Device		
									A070101	Failure to Capture		
									A070909	Oversensing		
									A0712	Pacing Problem		
									A072201	High Impedance		
							F2203	Imaging Required	A0203	Defective Device		
									A070101	Failure to Capture		
									A070909	Oversensing		
									A0712	Pacing Problem		
									A072201	High Impedance		
6068	2021	All other reportable incidents	60910	Coronary venous pacing lead	CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A0701	Capturing Problem		
									A0722	Impedance Problem		
									A072201	High Impedance		
							F26	No Health Consequences or Impact	A0701	Capturing Problem		
									A0722	Impedance Problem		

									A072201	High Impedance		
6069	2021	All other reportable incidents	47265	Dual-chamber implantable pacemaker, rate-responsive	CT141	Electrodes/Leads and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem		
							F26	No Health Consequences or Impact	A0705	Battery Problem		
					CT145	Cardiovascular devices	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem		
							F26	No Health Consequences or Impact	A0705	Battery Problem		
					CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem		
							F26	No Health Consequences or Impact	A0705	Battery Problem		
					CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem		
							F26	No Health Consequences or Impact	A0705	Battery Problem		
					CT233	Surgical	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem		
							F26	No Health Consequences or Impact	A0705	Battery Problem		
					CT2406	Implantable	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem		
							F26	No Health Consequences or Impact	A0705	Battery Problem		
					CT2407	Active implantable	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem		
							F26	No Health Consequences or Impact	A0705	Battery Problem		
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem		
							F26	No Health Consequences or Impact	A0705	Battery Problem		
					CT2727	Active implant devices/ancillaries	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem		
							F26	No Health Consequences or Impact	A0705	Battery Problem		
					CT2750	Implantable cardiac electrical stimulators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem		
							F26	No Health Consequences or Impact	A0705	Battery Problem		
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem		
							F26	No Health Consequences or Impact	A0705	Battery Problem		
					CT2985	Active	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem		
							F26	No Health Consequences or Impact	A0705	Battery Problem		
					CT3001	Implantable metal	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem		
							F26	No Health Consequences or Impact	A0705	Battery Problem		
					CT3002	Implantable synthetic polymer	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem		
							F26	No Health Consequences or Impact	A0705	Battery Problem		
					CT3006	Implantable inorganic material	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem		
							F26	No Health Consequences or Impact	A0705	Battery Problem		
					CT314	Electrically powered	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem		
							F26	No Health Consequences or Impact	A0705	Battery Problem		
					CT321	Long-term surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem		
							F26	No Health Consequences or Impact	A0705	Battery Problem		
					CT449	Generators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem		
							F26	No Health Consequences or Impact	A0705	Battery Problem		
					CT981	Single use	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem		
							F26	No Health Consequences or Impact	A0705	Battery Problem		
					CT983	Surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem		
							F26	No Health Consequences or Impact	A0705	Battery Problem		
6070	2021	All other reportable incidents	36241	Pacemaker/defibrillator lead	CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A072201	High Impedance		
							F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A072201	High Impedance		
							F26	No Health Consequences or Impact	A072201	High Impedance		
6071	2022	All other reportable incidents	47265	Dual-chamber implantable pacemaker, rate-responsive	CT141	Electrodes/Leads and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A0203	Defective Device	23/02/2022	23/02/2022
									A070504	Premature Discharge of Battery	23/02/2022	23/02/2022
							F08	Hospitalization or Prolonged Hospitalization	A0203	Defective Device	23/02/2022	23/02/2022
									A070504	Premature Discharge of Battery	23/02/2022	23/02/2022
							F1905	Device Revision or Replacement	A0203	Defective Device	23/02/2022	23/02/2022
									A070504	Premature Discharge of Battery	23/02/2022	23/02/2022
					CT145	Cardiovascular devices	E2403	No Clinical Signs, Symptoms or Conditions	A0203	Defective Device	23/02/2022	23/02/2022
									A070504	Premature Discharge of Battery	23/02/2022	23/02/2022
							F08	Hospitalization or Prolonged Hospitalization	A0203	Defective Device	23/02/2022	23/02/2022
									A070504	Premature Discharge of Battery	23/02/2022	23/02/2022
							F1905	Device Revision or Replacement	A0203	Defective Device	23/02/2022	23/02/2022
									A070504	Premature Discharge of Battery	23/02/2022	23/02/2022
					CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A0203	Defective Device	23/02/2022	23/02/2022
									A070504	Premature Discharge of Battery	23/02/2022	23/02/2022
							F08	Hospitalization or Prolonged Hospitalization	A0203	Defective Device	23/02/2022	23/02/2022
									A070504	Premature Discharge of Battery	23/02/2022	23/02/2022
							F1905	Device Revision or Replacement	A0203	Defective Device	23/02/2022	23/02/2022
									A070504	Premature Discharge of Battery	23/02/2022	23/02/2022
					CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A0203	Defective Device	23/02/2022	23/02/2022
									A070504	Premature Discharge of Battery	23/02/2022	23/02/2022
							F08	Hospitalization or Prolonged Hospitalization	A0203	Defective Device	23/02/2022	23/02/2022
									A070504	Premature Discharge of Battery	23/02/2022	23/02/2022
							F1905	Device Revision or Replacement	A0203	Defective Device	23/02/2022	23/02/2022
									A070504	Premature Discharge of Battery	23/02/2022	23/02/2022
					CT233	Surgical	E2403	No Clinical Signs, Symptoms or Conditions	A0203	Defective Device	23/02/2022	23/02/2022
									A070504	Premature Discharge of Battery	23/02/2022	23/02/2022
							F08	Hospitalization or Prolonged Hospitalization	A0203	Defective Device	23/02/2022	23/02/2022
									A070504	Premature Discharge of Battery	23/02/2022	23/02/2022
							F1905	Device Revision or Replacement	A0203	Defective Device	23/02/2022	23/02/2022
									A070504	Premature Discharge of Battery	23/02/2022	23/02/2022
					CT2406	Implantable	E2403	No Clinical Signs, Symptoms or Conditions	A0203	Defective Device	23/02/2022	23/02/2022
									A070504	Premature Discharge of Battery	23/02/2022	23/02/2022
							F08	Hospitalization or Prolonged Hospitalization	A0203	Defective Device	23/02/2022	23/02/2022
									A070504	Premature Discharge of Battery	23/02/2022	23/02/2022
							F1905	Device Revision or Replacement	A0203	Defective Device	23/02/2022	23/02/2022
									A070504	Premature Discharge of Battery	23/02/2022	23/02/2022
					CT2407	Active implantable	E2403	No Clinical Signs, Symptoms or Conditions	A0203	Defective Device	23/02/2022	23/02/2022
									A070504	Premature Discharge of Battery	23/02/2022	23/02/2022
							F08	Hospitalization or Prolonged Hospitalization	A0203	Defective Device	23/02/2022	23/02/2022
									A070504	Premature Discharge of Battery	23/02/2022	23/02/2022
							F1905	Device Revision or Replacement	A0203	Defective Device	23/02/2022	23/02/2022
									A070504	Premature Discharge of Battery	23/02/2022	23/02/2022
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A0203	Defective Device	23/02/2022	23/02/2022
									A070504	Premature Discharge of Battery	23/02/2022	23/02/2022

							F08	Hospitalization or Prolonged Hospitalization	A0203	Defective Device	23/02/2022	23/02/2022
									A070504	Premature Discharge of Battery	23/02/2022	23/02/2022
							F1905	Device Revision or Replacement	A0203	Defective Device	23/02/2022	23/02/2022
									A070504	Premature Discharge of Battery	23/02/2022	23/02/2022
					CT2727	Active implant devices/ancillaries	E2403	No Clinical Signs, Symptoms or Conditions	A0203	Defective Device	23/02/2022	23/02/2022
									A070504	Premature Discharge of Battery	23/02/2022	23/02/2022
							F08	Hospitalization or Prolonged Hospitalization	A0203	Defective Device	23/02/2022	23/02/2022
									A070504	Premature Discharge of Battery	23/02/2022	23/02/2022
							F1905	Device Revision or Replacement	A0203	Defective Device	23/02/2022	23/02/2022
									A070504	Premature Discharge of Battery	23/02/2022	23/02/2022
					CT2750	Implantable cardiac electrical stimulators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A0203	Defective Device	23/02/2022	23/02/2022
									A070504	Premature Discharge of Battery	23/02/2022	23/02/2022
							F08	Hospitalization or Prolonged Hospitalization	A0203	Defective Device	23/02/2022	23/02/2022
									A070504	Premature Discharge of Battery	23/02/2022	23/02/2022
							F1905	Device Revision or Replacement	A0203	Defective Device	23/02/2022	23/02/2022
									A070504	Premature Discharge of Battery	23/02/2022	23/02/2022
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A0203	Defective Device	23/02/2022	23/02/2022
									A070504	Premature Discharge of Battery	23/02/2022	23/02/2022
							F08	Hospitalization or Prolonged Hospitalization	A0203	Defective Device	23/02/2022	23/02/2022
									A070504	Premature Discharge of Battery	23/02/2022	23/02/2022
							F1905	Device Revision or Replacement	A0203	Defective Device	23/02/2022	23/02/2022
									A070504	Premature Discharge of Battery	23/02/2022	23/02/2022
					CT2985	Active	E2403	No Clinical Signs, Symptoms or Conditions	A0203	Defective Device	23/02/2022	23/02/2022
									A070504	Premature Discharge of Battery	23/02/2022	23/02/2022
							F08	Hospitalization or Prolonged Hospitalization	A0203	Defective Device	23/02/2022	23/02/2022
									A070504	Premature Discharge of Battery	23/02/2022	23/02/2022
							F1905	Device Revision or Replacement	A0203	Defective Device	23/02/2022	23/02/2022
									A070504	Premature Discharge of Battery	23/02/2022	23/02/2022
					CT3001	Implantable metal	E2403	No Clinical Signs, Symptoms or Conditions	A0203	Defective Device	23/02/2022	23/02/2022
									A070504	Premature Discharge of Battery	23/02/2022	23/02/2022
							F08	Hospitalization or Prolonged Hospitalization	A0203	Defective Device	23/02/2022	23/02/2022
									A070504	Premature Discharge of Battery	23/02/2022	23/02/2022
							F1905	Device Revision or Replacement	A0203	Defective Device	23/02/2022	23/02/2022
									A070504	Premature Discharge of Battery	23/02/2022	23/02/2022
					CT3002	Implantable synthetic polymer	E2403	No Clinical Signs, Symptoms or Conditions	A0203	Defective Device	23/02/2022	23/02/2022
									A070504	Premature Discharge of Battery	23/02/2022	23/02/2022
							F08	Hospitalization or Prolonged Hospitalization	A0203	Defective Device	23/02/2022	23/02/2022
									A070504	Premature Discharge of Battery	23/02/2022	23/02/2022
							F1905	Device Revision or Replacement	A0203	Defective Device	23/02/2022	23/02/2022
									A070504	Premature Discharge of Battery	23/02/2022	23/02/2022
					CT3006	Implantable inorganic material	E2403	No Clinical Signs, Symptoms or Conditions	A0203	Defective Device	23/02/2022	23/02/2022
									A070504	Premature Discharge of Battery	23/02/2022	23/02/2022
							F08	Hospitalization or Prolonged Hospitalization	A0203	Defective Device	23/02/2022	23/02/2022
									A070504	Premature Discharge of Battery	23/02/2022	23/02/2022
							F1905	Device Revision or Replacement	A0203	Defective Device	23/02/2022	23/02/2022
									A070504	Premature Discharge of Battery	23/02/2022	23/02/2022
					CT314	Electrically powered	E2403	No Clinical Signs, Symptoms or Conditions	A0203	Defective Device	23/02/2022	23/02/2022
									A070504	Premature Discharge of Battery	23/02/2022	23/02/2022
							F08	Hospitalization or Prolonged Hospitalization	A0203	Defective Device	23/02/2022	23/02/2022
									A070504	Premature Discharge of Battery	23/02/2022	23/02/2022
							F1905	Device Revision or Replacement	A0203	Defective Device	23/02/2022	23/02/2022
									A070504	Premature Discharge of Battery	23/02/2022	23/02/2022
					CT321	Long-term surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A0203	Defective Device	23/02/2022	23/02/2022
									A070504	Premature Discharge of Battery	23/02/2022	23/02/2022
							F08	Hospitalization or Prolonged Hospitalization	A0203	Defective Device	23/02/2022	23/02/2022
									A070504	Premature Discharge of Battery	23/02/2022	23/02/2022
							F1905	Device Revision or Replacement	A0203	Defective Device	23/02/2022	23/02/2022
									A070504	Premature Discharge of Battery	23/02/2022	23/02/2022
					CT449	Generators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A0203	Defective Device	23/02/2022	23/02/2022
									A070504	Premature Discharge of Battery	23/02/2022	23/02/2022
							F08	Hospitalization or Prolonged Hospitalization	A0203	Defective Device	23/02/2022	23/02/2022
									A070504	Premature Discharge of Battery	23/02/2022	23/02/2022
							F1905	Device Revision or Replacement	A0203	Defective Device	23/02/2022	23/02/2022
									A070504	Premature Discharge of Battery	23/02/2022	23/02/2022
					CT981	Single use	E2403	No Clinical Signs, Symptoms or Conditions	A0203	Defective Device	23/02/2022	23/02/2022
									A070504	Premature Discharge of Battery	23/02/2022	23/02/2022
							F08	Hospitalization or Prolonged Hospitalization	A0203	Defective Device	23/02/2022	23/02/2022
									A070504	Premature Discharge of Battery	23/02/2022	23/02/2022
							F1905	Device Revision or Replacement	A0203	Defective Device	23/02/2022	23/02/2022
									A070504	Premature Discharge of Battery	23/02/2022	23/02/2022
					CT983	Surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A0203	Defective Device	23/02/2022	23/02/2022
									A070504	Premature Discharge of Battery	23/02/2022	23/02/2022
							F08	Hospitalization or Prolonged Hospitalization	A0203	Defective Device	23/02/2022	23/02/2022
									A070504	Premature Discharge of Battery	23/02/2022	23/02/2022
							F1905	Device Revision or Replacement	A0203	Defective Device	23/02/2022	23/02/2022
									A070504	Premature Discharge of Battery	23/02/2022	23/02/2022
6072	2021	All other reportable incidents	47270	Cardiac resynchronization therapy implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E2104	Electric Shock	A070909	Oversensing		
									A0712	Pacing Problem		
									A071302	Inappropriate/Inadequate Shock/Stimulation		
							E2403	No Clinical Signs, Symptoms or Conditions	A070909	Oversensing		
									A0712	Pacing Problem		
									A071302	Inappropriate/Inadequate Shock/Stimulation		
							F26	No Health Consequences or Impact	A070909	Oversensing		
									A0712	Pacing Problem		
									A071302	Inappropriate/Inadequate Shock/Stimulation		
6073	2021	All other reportable incidents	35852	Single-chamber implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem	30/07/2021	30/07/2021
									A070504	Premature Discharge of Battery	30/07/2021	30/07/2021
							F08	Hospitalization or Prolonged Hospitalization	A0705	Battery Problem	30/07/2021	30/07/2021

									A070504	Premature Discharge of Battery	30/07/2021	30/07/2021
							F1905	Device Revision or Replacement	A0705	Battery Problem	30/07/2021	30/07/2021
									A070504	Premature Discharge of Battery	30/07/2021	30/07/2021
6074	2022	All other reportable incidents	47265	Dual-chamber implantable pacemaker, rate-responsive	CT141	Electrodes/Leads and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	04/11/2021	04/11/2021
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	04/11/2021	04/11/2021
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	04/11/2021	04/11/2021
					CT145	Cardiovascular devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	04/11/2021	04/11/2021
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	04/11/2021	04/11/2021
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	04/11/2021	04/11/2021
					CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	04/11/2021	04/11/2021
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	04/11/2021	04/11/2021
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	04/11/2021	04/11/2021
					CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	04/11/2021	04/11/2021
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	04/11/2021	04/11/2021
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	04/11/2021	04/11/2021
					CT233	Surgical	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	04/11/2021	04/11/2021
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	04/11/2021	04/11/2021
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	04/11/2021	04/11/2021
					CT2406	Implantable	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	04/11/2021	04/11/2021
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	04/11/2021	04/11/2021
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	04/11/2021	04/11/2021
					CT2407	Active implantable	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	04/11/2021	04/11/2021
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	04/11/2021	04/11/2021
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	04/11/2021	04/11/2021
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	04/11/2021	04/11/2021
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	04/11/2021	04/11/2021
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	04/11/2021	04/11/2021
					CT2727	Active implant devices/ancillaries	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	04/11/2021	04/11/2021
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	04/11/2021	04/11/2021
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	04/11/2021	04/11/2021
					CT2750	Implantable cardiac electrical stimulators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	04/11/2021	04/11/2021
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	04/11/2021	04/11/2021
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	04/11/2021	04/11/2021
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	04/11/2021	04/11/2021
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	04/11/2021	04/11/2021
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	04/11/2021	04/11/2021
					CT2985	Active	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	04/11/2021	04/11/2021
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	04/11/2021	04/11/2021
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	04/11/2021	04/11/2021
					CT3001	Implantable metal	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	04/11/2021	04/11/2021
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	04/11/2021	04/11/2021
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	04/11/2021	04/11/2021
					CT3002	Implantable synthetic polymer	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	04/11/2021	04/11/2021
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	04/11/2021	04/11/2021
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	04/11/2021	04/11/2021
					CT3006	Implantable inorganic material	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	04/11/2021	04/11/2021
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	04/11/2021	04/11/2021
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	04/11/2021	04/11/2021
					CT314	Electrically powered	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	04/11/2021	04/11/2021
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	04/11/2021	04/11/2021
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	04/11/2021	04/11/2021
					CT321	Long-term surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	04/11/2021	04/11/2021
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	04/11/2021	04/11/2021
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	04/11/2021	04/11/2021
					CT449	Generators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	04/11/2021	04/11/2021
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	04/11/2021	04/11/2021
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	04/11/2021	04/11/2021
					CT981	Single use	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	04/11/2021	04/11/2021
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	04/11/2021	04/11/2021
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	04/11/2021	04/11/2021
					CT983	Surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	04/11/2021	04/11/2021
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	04/11/2021	04/11/2021
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	04/11/2021	04/11/2021
6075	2022	All other reportable incidents	47265	Dual-chamber implantable pacemaker, rate-responsive	CT141	Electrodes/Leads and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	18/03/2022	18/03/2022
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	18/03/2022	18/03/2022
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	18/03/2022	18/03/2022
					CT145	Cardiovascular devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	18/03/2022	18/03/2022
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	18/03/2022	18/03/2022
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	18/03/2022	18/03/2022
					CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	18/03/2022	18/03/2022
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	18/03/2022	18/03/2022
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	18/03/2022	18/03/2022
					CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	18/03/2022	18/03/2022
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	18/03/2022	18/03/2022
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	18/03/2022	18/03/2022
					CT233	Surgical	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	18/03/2022	18/03/2022
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	18/03/2022	18/03/2022
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	18/03/2022	18/03/2022
					CT2406	Implantable	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	18/03/2022	18/03/2022
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	18/03/2022	18/03/2022
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	18/03/2022	18/03/2022
					CT2407	Active implantable	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	18/03/2022	18/03/2022
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	18/03/2022	18/03/2022
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	18/03/2022	18/03/2022
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	18/03/2022	18/03/2022
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	18/03/2022	18/03/2022
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	18/03/2022	18/03/2022
					CT2727	Active implant devices/ancillaries	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	18/03/2022	18/03/2022

							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	18/03/2022	18/03/2022
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	18/03/2022	18/03/2022
					CT2750	Implantable cardiac electrical stimulators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	18/03/2022	18/03/2022
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	18/03/2022	18/03/2022
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	18/03/2022	18/03/2022
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	18/03/2022	18/03/2022
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	18/03/2022	18/03/2022
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	18/03/2022	18/03/2022
					CT2985	Active	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	18/03/2022	18/03/2022
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	18/03/2022	18/03/2022
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	18/03/2022	18/03/2022
					CT3001	Implantable metal	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	18/03/2022	18/03/2022
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	18/03/2022	18/03/2022
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	18/03/2022	18/03/2022
					CT3002	Implantable synthetic polymer	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	18/03/2022	18/03/2022
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	18/03/2022	18/03/2022
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	18/03/2022	18/03/2022
					CT3006	Implantable inorganic material	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	18/03/2022	18/03/2022
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	18/03/2022	18/03/2022
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	18/03/2022	18/03/2022
					CT314	Electrically powered	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	18/03/2022	18/03/2022
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	18/03/2022	18/03/2022
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	18/03/2022	18/03/2022
					CT321	Long-term surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	18/03/2022	18/03/2022
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	18/03/2022	18/03/2022
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	18/03/2022	18/03/2022
					CT449	Generators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	18/03/2022	18/03/2022
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	18/03/2022	18/03/2022
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	18/03/2022	18/03/2022
					CT981	Single use	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	18/03/2022	18/03/2022
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	18/03/2022	18/03/2022
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	18/03/2022	18/03/2022
					CT983	Surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	18/03/2022	18/03/2022
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	18/03/2022	18/03/2022
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	18/03/2022	18/03/2022
6076	2022	All other reportable incidents	47270	Cardiac resynchronization therapy implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A072201	High Impedance		
									A0908	Incorrect, Inadequate or Imprecise Result or Readings		
							F26	No Health Consequences or Impact	A072201	High Impedance		
									A0908	Incorrect, Inadequate or Imprecise Result or Readings		
6077	2022	All other reportable incidents	17882	Physiologic-monitoring defibrillation system	CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070803	Failure to Power Up		
							F27	Problem Identified During Non-Clinical Procedure	A070803	Failure to Power Up		
6078	2021	All other reportable incidents	47265	Dual-chamber implantable pacemaker, rate-responsive	CT141	Electrodes/Leads and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT145	Cardiovascular devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT233	Surgical	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT2406	Implantable	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT2407	Active implantable	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT2727	Active implant devices/ancillaries	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT2750	Implantable cardiac electrical stimulators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT2985	Active	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT3001	Implantable metal	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT3002	Implantable synthetic polymer	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT3006	Implantable inorganic material	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT314	Electrically powered	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT321	Long-term surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT449	Generators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT981	Single use	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT983	Surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
6079	2021	All other reportable incidents	35223	Endocardial/interventricular septal pacing lead	CT1684	Cardiac pacemakers and associated devices	E2311	Discomfort	A040101	Fracture		
									A0701	Capturing Problem		
									A0709	Device Sensing Problem		
							E2403	No Clinical Signs, Symptoms or Conditions	A040101	Fracture		
									A0701	Capturing Problem		

									A0709	Device Sensing Problem		
							F23	Unexpected Medical Intervention	A040101	Fracture		
									A0701	Capturing Problem		
									A0709	Device Sensing Problem		
6080	2022	All other reportable incidents	47265	Dual-chamber implantable pacemaker, rate-responsive	CT141	Electrodes/Leads and associated devices	E011903	Syncope/Fainting	A070504	Premature Discharge of Battery	28/02/2022	28/02/2022
									A071102	Failure to Interrogate	28/02/2022	28/02/2022
									A071204	No Pacing	28/02/2022	28/02/2022
							E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	28/02/2022	28/02/2022
									A071102	Failure to Interrogate	28/02/2022	28/02/2022
									A071204	No Pacing	28/02/2022	28/02/2022
							F1901	Additional Surgery	A070504	Premature Discharge of Battery	28/02/2022	28/02/2022
									A071102	Failure to Interrogate	28/02/2022	28/02/2022
									A071204	No Pacing	28/02/2022	28/02/2022
					CT145	Cardiovascular devices	E011903	Syncope/Fainting	A070504	Premature Discharge of Battery	28/02/2022	28/02/2022
									A071102	Failure to Interrogate	28/02/2022	28/02/2022
									A071204	No Pacing	28/02/2022	28/02/2022
							E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	28/02/2022	28/02/2022
									A071102	Failure to Interrogate	28/02/2022	28/02/2022
									A071204	No Pacing	28/02/2022	28/02/2022
							F1901	Additional Surgery	A070504	Premature Discharge of Battery	28/02/2022	28/02/2022
									A071102	Failure to Interrogate	28/02/2022	28/02/2022
									A071204	No Pacing	28/02/2022	28/02/2022
					CT1684	Cardiac pacemakers and associated devices	E011903	Syncope/Fainting	A070504	Premature Discharge of Battery	28/02/2022	28/02/2022
									A071102	Failure to Interrogate	28/02/2022	28/02/2022
									A071204	No Pacing	28/02/2022	28/02/2022
							E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	28/02/2022	28/02/2022
									A071102	Failure to Interrogate	28/02/2022	28/02/2022
									A071204	No Pacing	28/02/2022	28/02/2022
							F1901	Additional Surgery	A070504	Premature Discharge of Battery	28/02/2022	28/02/2022
									A071102	Failure to Interrogate	28/02/2022	28/02/2022
									A071204	No Pacing	28/02/2022	28/02/2022
					CT1685	Cardiac pacemakers	E011903	Syncope/Fainting	A070504	Premature Discharge of Battery	28/02/2022	28/02/2022
									A071102	Failure to Interrogate	28/02/2022	28/02/2022
									A071204	No Pacing	28/02/2022	28/02/2022
							E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	28/02/2022	28/02/2022
									A071102	Failure to Interrogate	28/02/2022	28/02/2022
									A071204	No Pacing	28/02/2022	28/02/2022
							F1901	Additional Surgery	A070504	Premature Discharge of Battery	28/02/2022	28/02/2022
									A071102	Failure to Interrogate	28/02/2022	28/02/2022
									A071204	No Pacing	28/02/2022	28/02/2022
					CT233	Surgical	E011903	Syncope/Fainting	A070504	Premature Discharge of Battery	28/02/2022	28/02/2022
									A071102	Failure to Interrogate	28/02/2022	28/02/2022
									A071204	No Pacing	28/02/2022	28/02/2022
							E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	28/02/2022	28/02/2022
									A071102	Failure to Interrogate	28/02/2022	28/02/2022
									A071204	No Pacing	28/02/2022	28/02/2022
							F1901	Additional Surgery	A070504	Premature Discharge of Battery	28/02/2022	28/02/2022
									A071102	Failure to Interrogate	28/02/2022	28/02/2022
									A071204	No Pacing	28/02/2022	28/02/2022
					CT2406	Implantable	E011903	Syncope/Fainting	A070504	Premature Discharge of Battery	28/02/2022	28/02/2022
									A071102	Failure to Interrogate	28/02/2022	28/02/2022
									A071204	No Pacing	28/02/2022	28/02/2022
							E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	28/02/2022	28/02/2022
									A071102	Failure to Interrogate	28/02/2022	28/02/2022
									A071204	No Pacing	28/02/2022	28/02/2022
							F1901	Additional Surgery	A070504	Premature Discharge of Battery	28/02/2022	28/02/2022
									A071102	Failure to Interrogate	28/02/2022	28/02/2022
									A071204	No Pacing	28/02/2022	28/02/2022
					CT2407	Active implantable	E011903	Syncope/Fainting	A070504	Premature Discharge of Battery	28/02/2022	28/02/2022
									A071102	Failure to Interrogate	28/02/2022	28/02/2022
									A071204	No Pacing	28/02/2022	28/02/2022
							E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	28/02/2022	28/02/2022
									A071102	Failure to Interrogate	28/02/2022	28/02/2022
									A071204	No Pacing	28/02/2022	28/02/2022
							F1901	Additional Surgery	A070504	Premature Discharge of Battery	28/02/2022	28/02/2022
									A071102	Failure to Interrogate	28/02/2022	28/02/2022
									A071204	No Pacing	28/02/2022	28/02/2022
					CT257	Implantable cardiac pacemakers	E011903	Syncope/Fainting	A070504	Premature Discharge of Battery	28/02/2022	28/02/2022
									A071102	Failure to Interrogate	28/02/2022	28/02/2022
									A071204	No Pacing	28/02/2022	28/02/2022
							E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	28/02/2022	28/02/2022
									A071102	Failure to Interrogate	28/02/2022	28/02/2022
									A071204	No Pacing	28/02/2022	28/02/2022
							F1901	Additional Surgery	A070504	Premature Discharge of Battery	28/02/2022	28/02/2022
									A071102	Failure to Interrogate	28/02/2022	28/02/2022
									A071204	No Pacing	28/02/2022	28/02/2022
					CT2727	Active implant devices/ancillaries	E011903	Syncope/Fainting	A070504	Premature Discharge of Battery	28/02/2022	28/02/2022
									A071102	Failure to Interrogate	28/02/2022	28/02/2022
									A071204	No Pacing	28/02/2022	28/02/2022
							E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	28/02/2022	28/02/2022
									A071102	Failure to Interrogate	28/02/2022	28/02/2022
									A071204	No Pacing	28/02/2022	28/02/2022
							F1901	Additional Surgery	A070504	Premature Discharge of Battery	28/02/2022	28/02/2022
									A071102	Failure to Interrogate	28/02/2022	28/02/2022
									A071204	No Pacing	28/02/2022	28/02/2022
					CT2750	Implantable cardiac electrical stimulators and associated devices	E011903	Syncope/Fainting	A070504	Premature Discharge of Battery	28/02/2022	28/02/2022
									A071102	Failure to Interrogate	28/02/2022	28/02/2022
									A071204	No Pacing	28/02/2022	28/02/2022

						E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	28/02/2022	28/02/2022
								A071102	Failure to Interrogate	28/02/2022	28/02/2022
								A071204	No Pacing	28/02/2022	28/02/2022
						F1901	Additional Surgery	A070504	Premature Discharge of Battery	28/02/2022	28/02/2022
								A071102	Failure to Interrogate	28/02/2022	28/02/2022
								A071204	No Pacing	28/02/2022	28/02/2022
					CT2866	Implantable cardiac electrical stimulators	E011903	Syncope/Fainting	A070504	Premature Discharge of Battery	28/02/2022	28/02/2022
									A071102	Failure to Interrogate	28/02/2022	28/02/2022
									A071204	No Pacing	28/02/2022	28/02/2022
							E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	28/02/2022	28/02/2022
									A071102	Failure to Interrogate	28/02/2022	28/02/2022
									A071204	No Pacing	28/02/2022	28/02/2022
							F1901	Additional Surgery	A070504	Premature Discharge of Battery	28/02/2022	28/02/2022
									A071102	Failure to Interrogate	28/02/2022	28/02/2022
									A071204	No Pacing	28/02/2022	28/02/2022
					CT2985	Active	E011903	Syncope/Fainting	A070504	Premature Discharge of Battery	28/02/2022	28/02/2022
									A071102	Failure to Interrogate	28/02/2022	28/02/2022
									A071204	No Pacing	28/02/2022	28/02/2022
							E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	28/02/2022	28/02/2022
									A071102	Failure to Interrogate	28/02/2022	28/02/2022
									A071204	No Pacing	28/02/2022	28/02/2022
							F1901	Additional Surgery	A070504	Premature Discharge of Battery	28/02/2022	28/02/2022
									A071102	Failure to Interrogate	28/02/2022	28/02/2022
									A071204	No Pacing	28/02/2022	28/02/2022
					CT3001	Implantable metal	E011903	Syncope/Fainting	A070504	Premature Discharge of Battery	28/02/2022	28/02/2022
									A071102	Failure to Interrogate	28/02/2022	28/02/2022
									A071204	No Pacing	28/02/2022	28/02/2022
							E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	28/02/2022	28/02/2022
									A071102	Failure to Interrogate	28/02/2022	28/02/2022
									A071204	No Pacing	28/02/2022	28/02/2022
							F1901	Additional Surgery	A070504	Premature Discharge of Battery	28/02/2022	28/02/2022
									A071102	Failure to Interrogate	28/02/2022	28/02/2022
									A071204	No Pacing	28/02/2022	28/02/2022
					CT3002	Implantable synthetic polymer	E011903	Syncope/Fainting	A070504	Premature Discharge of Battery	28/02/2022	28/02/2022
									A071102	Failure to Interrogate	28/02/2022	28/02/2022
									A071204	No Pacing	28/02/2022	28/02/2022
							E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	28/02/2022	28/02/2022
									A071102	Failure to Interrogate	28/02/2022	28/02/2022
									A071204	No Pacing	28/02/2022	28/02/2022
							F1901	Additional Surgery	A070504	Premature Discharge of Battery	28/02/2022	28/02/2022
									A071102	Failure to Interrogate	28/02/2022	28/02/2022
									A071204	No Pacing	28/02/2022	28/02/2022
					CT3006	Implantable inorganic material	E011903	Syncope/Fainting	A070504	Premature Discharge of Battery	28/02/2022	28/02/2022
									A071102	Failure to Interrogate	28/02/2022	28/02/2022
									A071204	No Pacing	28/02/2022	28/02/2022
							E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	28/02/2022	28/02/2022
									A071102	Failure to Interrogate	28/02/2022	28/02/2022
									A071204	No Pacing	28/02/2022	28/02/2022
							F1901	Additional Surgery	A070504	Premature Discharge of Battery	28/02/2022	28/02/2022
									A071102	Failure to Interrogate	28/02/2022	28/02/2022
									A071204	No Pacing	28/02/2022	28/02/2022
					CT314	Electrically powered	E011903	Syncope/Fainting	A070504	Premature Discharge of Battery	28/02/2022	28/02/2022
									A071102	Failure to Interrogate	28/02/2022	28/02/2022
									A071204	No Pacing	28/02/2022	28/02/2022
							E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	28/02/2022	28/02/2022
									A071102	Failure to Interrogate	28/02/2022	28/02/2022
									A071204	No Pacing	28/02/2022	28/02/2022
							F1901	Additional Surgery	A070504	Premature Discharge of Battery	28/02/2022	28/02/2022
									A071102	Failure to Interrogate	28/02/2022	28/02/2022
									A071204	No Pacing	28/02/2022	28/02/2022
					CT321	Long-term surgical invasive	E011903	Syncope/Fainting	A070504	Premature Discharge of Battery	28/02/2022	28/02/2022
									A071102	Failure to Interrogate	28/02/2022	28/02/2022
									A071204	No Pacing	28/02/2022	28/02/2022
							E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	28/02/2022	28/02/2022
									A071102	Failure to Interrogate	28/02/2022	28/02/2022
									A071204	No Pacing	28/02/2022	28/02/2022
							F1901	Additional Surgery	A070504	Premature Discharge of Battery	28/02/2022	28/02/2022
									A071102	Failure to Interrogate	28/02/2022	28/02/2022
									A071204	No Pacing	28/02/2022	28/02/2022
					CT449	Generators and associated devices	E011903	Syncope/Fainting	A070504	Premature Discharge of Battery	28/02/2022	28/02/2022
									A071102	Failure to Interrogate	28/02/2022	28/02/2022
									A071204	No Pacing	28/02/2022	28/02/2022
							E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	28/02/2022	28/02/2022
									A071102	Failure to Interrogate	28/02/2022	28/02/2022
									A071204	No Pacing	28/02/2022	28/02/2022
							F1901	Additional Surgery	A070504	Premature Discharge of Battery	28/02/2022	28/02/2022
									A071102	Failure to Interrogate	28/02/2022	28/02/2022
									A071204	No Pacing	28/02/2022	28/02/2022
					CT981	Single use	E011903	Syncope/Fainting	A070504	Premature Discharge of Battery	28/02/2022	28/02/2022
									A071102	Failure to Interrogate	28/02/2022	28/02/2022
									A071204	No Pacing	28/02/2022	28/02/2022
							E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	28/02/2022	28/02/2022
									A071102	Failure to Interrogate	28/02/2022	28/02/2022
									A071204	No Pacing	28/02/2022	28/02/2022
							F1901	Additional Surgery	A070504	Premature Discharge of Battery	28/02/2022	28/02/2022
									A071102	Failure to Interrogate	28/02/2022	28/02/2022
									A071204	No Pacing	28/02/2022	28/02/2022
					CT983	Surgical invasive	E011903	Syncope/Fainting	A070504	Premature Discharge of Battery	28/02/2022	28/02/2022

								A071102	Failure to Interrogate	28/02/2022	28/02/2022
								A071204	No Pacing	28/02/2022	28/02/2022
						E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	28/02/2022	28/02/2022
								A071102	Failure to Interrogate	28/02/2022	28/02/2022
								A071204	No Pacing	28/02/2022	28/02/2022
						F1901	Additional Surgery	A070504	Premature Discharge of Battery	28/02/2022	28/02/2022
								A071102	Failure to Interrogate	28/02/2022	28/02/2022
								A071204	No Pacing	28/02/2022	28/02/2022
6081	2021	All other reportable incidents	47263	Cardiac resynchronization therapy implantable pacemaker	CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A0908	Incorrect, Inadequate or Imprecise Result or Readings	13/07/2021	13/07/2021
							F05	Delay to Treatment/ Therapy	A0908	Incorrect, Inadequate or Imprecise Result or Readings	13/07/2021	13/07/2021
							F08	Hospitalization or Prolonged Hospitalization	A0908	Incorrect, Inadequate or Imprecise Result or Readings	13/07/2021	13/07/2021
							F1905	Device Revision or Replacement	A0908	Incorrect, Inadequate or Imprecise Result or Readings	13/07/2021	13/07/2021
					CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A0908	Incorrect, Inadequate or Imprecise Result or Readings	13/07/2021	13/07/2021
							F05	Delay to Treatment/ Therapy	A0908	Incorrect, Inadequate or Imprecise Result or Readings	13/07/2021	13/07/2021
							F08	Hospitalization or Prolonged Hospitalization	A0908	Incorrect, Inadequate or Imprecise Result or Readings	13/07/2021	13/07/2021
							F1905	Device Revision or Replacement	A0908	Incorrect, Inadequate or Imprecise Result or Readings	13/07/2021	13/07/2021
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A0908	Incorrect, Inadequate or Imprecise Result or Readings	13/07/2021	13/07/2021
							F05	Delay to Treatment/ Therapy	A0908	Incorrect, Inadequate or Imprecise Result or Readings	13/07/2021	13/07/2021
							F08	Hospitalization or Prolonged Hospitalization	A0908	Incorrect, Inadequate or Imprecise Result or Readings	13/07/2021	13/07/2021
							F1905	Device Revision or Replacement	A0908	Incorrect, Inadequate or Imprecise Result or Readings	13/07/2021	13/07/2021
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A0908	Incorrect, Inadequate or Imprecise Result or Readings	13/07/2021	13/07/2021
							F05	Delay to Treatment/ Therapy	A0908	Incorrect, Inadequate or Imprecise Result or Readings	13/07/2021	13/07/2021
							F08	Hospitalization or Prolonged Hospitalization	A0908	Incorrect, Inadequate or Imprecise Result or Readings	13/07/2021	13/07/2021
							F1905	Device Revision or Replacement	A0908	Incorrect, Inadequate or Imprecise Result or Readings	13/07/2021	13/07/2021
6082	2021	All other reportable incidents	35852	Single-chamber implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	13/10/2021	13/10/2021
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	13/10/2021	13/10/2021
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	13/10/2021	13/10/2021
6083	2021	All other reportable incidents	47265	Dual-chamber implantable pacemaker, rate-responsive	CT141	Electrodes/Leads and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT145	Cardiovascular devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT233	Surgical	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT2406	Implantable	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT2407	Active implantable	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT2727	Active implant devices/ancillaries	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT2750	Implantable cardiac electrical stimulators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT2985	Active	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT3001	Implantable metal	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT3002	Implantable synthetic polymer	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT3006	Implantable inorganic material	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT314	Electrically powered	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT321	Long-term surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT449	Generators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT981	Single use	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT983	Surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
6084	2022	All other reportable incidents	17882	Physiologic-monitoring defibrillation system	CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A1603	Failure of Device to Self-Test		
							F26	No Health Consequences or Impact	A1603	Failure of Device to Self-Test		
6085	2021	All other reportable incidents	35852	Single-chamber implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	09/07/2021	09/07/2021
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	09/07/2021	09/07/2021
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	09/07/2021	09/07/2021
6086	2022	All other reportable incidents	36241	Pacemaker/defibrillator lead	CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A072201	High Impedance		
							F26	No Health Consequences or Impact	A072201	High Impedance		
6087	2022	All other reportable incidents	35852	Single-chamber implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E2104	Electric Shock	A070909	Oversensing		
									A071302	Inappropriate/Inadequate Shock/Stimulation		

							E2403	No Clinical Signs, Symptoms or Conditions	A070909	Oversensing		
									A071302	Inappropriate/Inadequate Shock/Stimulation		
							F08	Hospitalization or Prolonged Hospitalization	A070909	Oversensing		
									A071302	Inappropriate/Inadequate Shock/Stimulation		
							F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A070909	Oversensing		
									A071302	Inappropriate/Inadequate Shock/Stimulation		
							F1905	Device Revision or Replacement	A070909	Oversensing		
									A071302	Inappropriate/Inadequate Shock/Stimulation		
6088	2021	All other reportable incidents	35852	Single-chamber implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem	09/09/2021	09/09/2021
									A070504	Premature Discharge of Battery	09/09/2021	09/09/2021
							F08	Hospitalization or Prolonged Hospitalization	A0705	Battery Problem	09/09/2021	09/09/2021
									A070504	Premature Discharge of Battery	09/09/2021	09/09/2021
							F1905	Device Revision or Replacement	A0705	Battery Problem	09/09/2021	09/09/2021
									A070504	Premature Discharge of Battery	09/09/2021	09/09/2021
6089	2022	All other reportable incidents	47270	Cardiac resynchronization therapy implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E060109	Tachycardia	A0712	Pacing Problem		
							E2403	No Clinical Signs, Symptoms or Conditions	A0712	Pacing Problem		
							F12	Serious Injury/ Illness/ Impairment	A0712	Pacing Problem		
							F23	Unexpected Medical Intervention	A0712	Pacing Problem		
6090	2021	Death	17882	Physiologic-monitoring defibrillation system	CT1684	Cardiac pacemakers and associated devices	E0602	Cardiac Arrest	A090811	Unable to Obtain Readings		
							F02	Death	A090811	Unable to Obtain Readings		
6091	2022	All other reportable incidents	36241	Pacemaker/defibrillator lead	CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A072201	High Impedance		
							F26	No Health Consequences or Impact	A072201	High Impedance		
6092	2021	All other reportable incidents	35223	Endocardial/interventricular septal pacing lead	CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A0401	Break		
									A07	Electrical /Electronic Property Problem		
									A070909	Oversensing		
									A072202	Low Impedance		
									A090801	Signal Artifact/Noise		
							F26	No Health Consequences or Impact	A0401	Break		
									A07	Electrical /Electronic Property Problem		
									A070909	Oversensing		
									A072202	Low Impedance		
									A090801	Signal Artifact/Noise		
6093	2021	All other reportable incidents	35223	Endocardial/interventricular septal pacing lead	CT1684	Cardiac pacemakers and associated devices	E0112	Dizziness	A0401	Break		
									A070909	Oversensing		
									A0712	Pacing Problem		
									A090801	Signal Artifact/Noise		
							E2403	No Clinical Signs, Symptoms or Conditions	A0401	Break		
									A070909	Oversensing		
									A0712	Pacing Problem		
									A090801	Signal Artifact/Noise		
							F11	Non-Serious Injury/ Illness/ Impairment	A0401	Break		
									A070909	Oversensing		
									A0712	Pacing Problem		
									A090801	Signal Artifact/Noise		
							F26	No Health Consequences or Impact	A0401	Break		
									A070909	Oversensing		
									A0712	Pacing Problem		
									A090801	Signal Artifact/Noise		
6094	2021	All other reportable incidents	35223	Endocardial/interventricular septal pacing lead	CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070909	Oversensing		
									A072001	Electromagnetic Interference		
									A0722	Impedance Problem		
									A072201	High Impedance		
									A090801	Signal Artifact/Noise		
							F26	No Health Consequences or Impact	A070909	Oversensing		
									A072001	Electromagnetic Interference		
									A0722	Impedance Problem		
									A072201	High Impedance		
									A090801	Signal Artifact/Noise		
6095	2022	All other reportable incidents	47265	Dual-chamber implantable pacemaker, rate-responsive	CT141	Electrodes/Leads and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	14/04/2022	14/04/2022
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	14/04/2022	14/04/2022
							F1903	Device Explantation	A070504	Premature Discharge of Battery	14/04/2022	14/04/2022
					CT145	Cardiovascular devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	14/04/2022	14/04/2022
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	14/04/2022	14/04/2022
							F1903	Device Explantation	A070504	Premature Discharge of Battery	14/04/2022	14/04/2022
					CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	14/04/2022	14/04/2022
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	14/04/2022	14/04/2022
							F1903	Device Explantation	A070504	Premature Discharge of Battery	14/04/2022	14/04/2022
					CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	14/04/2022	14/04/2022
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	14/04/2022	14/04/2022
							F1903	Device Explantation	A070504	Premature Discharge of Battery	14/04/2022	14/04/2022
					CT233	Surgical	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	14/04/2022	14/04/2022
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	14/04/2022	14/04/2022
							F1903	Device Explantation	A070504	Premature Discharge of Battery	14/04/2022	14/04/2022
					CT2406	Implantable	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	14/04/2022	14/04/2022
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	14/04/2022	14/04/2022
							F1903	Device Explantation	A070504	Premature Discharge of Battery	14/04/2022	14/04/2022
					CT2407	Active implantable	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	14/04/2022	14/04/2022
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	14/04/2022	14/04/2022
							F1903	Device Explantation	A070504	Premature Discharge of Battery	14/04/2022	14/04/2022
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	14/04/2022	14/04/2022
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	14/04/2022	14/04/2022
							F1903	Device Explantation	A070504	Premature Discharge of Battery	14/04/2022	14/04/2022
					CT2727	Active implant devices/ancillaries	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	14/04/2022	14/04/2022
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	14/04/2022	14/04/2022
							F1903	Device Explantation	A070504	Premature Discharge of Battery	14/04/2022	14/04/2022
					CT2750	Implantable cardiac electrical stimulators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	14/04/2022	14/04/2022
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	14/04/2022	14/04/2022
							F1903	Device Explantation	A070504	Premature Discharge of Battery	14/04/2022	14/04/2022

					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	14/04/2022	14/04/2022
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	14/04/2022	14/04/2022
							F1903	Device Explantation	A070504	Premature Discharge of Battery	14/04/2022	14/04/2022
					CT2985	Active	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	14/04/2022	14/04/2022
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	14/04/2022	14/04/2022
							F1903	Device Explantation	A070504	Premature Discharge of Battery	14/04/2022	14/04/2022
					CT3001	Implantable metal	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	14/04/2022	14/04/2022
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	14/04/2022	14/04/2022
							F1903	Device Explantation	A070504	Premature Discharge of Battery	14/04/2022	14/04/2022
					CT3002	Implantable synthetic polymer	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	14/04/2022	14/04/2022
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	14/04/2022	14/04/2022
							F1903	Device Explantation	A070504	Premature Discharge of Battery	14/04/2022	14/04/2022
					CT3006	Implantable inorganic material	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	14/04/2022	14/04/2022
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	14/04/2022	14/04/2022
							F1903	Device Explantation	A070504	Premature Discharge of Battery	14/04/2022	14/04/2022
					CT314	Electrically powered	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	14/04/2022	14/04/2022
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	14/04/2022	14/04/2022
							F1903	Device Explantation	A070504	Premature Discharge of Battery	14/04/2022	14/04/2022
					CT321	Long-term surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	14/04/2022	14/04/2022
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	14/04/2022	14/04/2022
							F1903	Device Explantation	A070504	Premature Discharge of Battery	14/04/2022	14/04/2022
					CT449	Generators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	14/04/2022	14/04/2022
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	14/04/2022	14/04/2022
							F1903	Device Explantation	A070504	Premature Discharge of Battery	14/04/2022	14/04/2022
					CT981	Single use	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	14/04/2022	14/04/2022
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	14/04/2022	14/04/2022
							F1903	Device Explantation	A070504	Premature Discharge of Battery	14/04/2022	14/04/2022
					CT983	Surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	14/04/2022	14/04/2022
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	14/04/2022	14/04/2022
							F1903	Device Explantation	A070504	Premature Discharge of Battery	14/04/2022	14/04/2022
6096	2022	All other reportable incidents	47265	Dual-chamber implantable pacemaker, rate-responsive	CT141	Electrodes/Leads and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	25/02/2022	25/02/2022
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	25/02/2022	25/02/2022
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	25/02/2022	25/02/2022
					CT145	Cardiovascular devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	25/02/2022	25/02/2022
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	25/02/2022	25/02/2022
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	25/02/2022	25/02/2022
					CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	25/02/2022	25/02/2022
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	25/02/2022	25/02/2022
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	25/02/2022	25/02/2022
					CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	25/02/2022	25/02/2022
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	25/02/2022	25/02/2022
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	25/02/2022	25/02/2022
					CT233	Surgical	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	25/02/2022	25/02/2022
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	25/02/2022	25/02/2022
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	25/02/2022	25/02/2022
					CT2406	Implantable	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	25/02/2022	25/02/2022
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	25/02/2022	25/02/2022
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	25/02/2022	25/02/2022
					CT2407	Active implantable	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	25/02/2022	25/02/2022
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	25/02/2022	25/02/2022
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	25/02/2022	25/02/2022
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	25/02/2022	25/02/2022
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	25/02/2022	25/02/2022
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	25/02/2022	25/02/2022
					CT2727	Active implant devices/ancillaries	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	25/02/2022	25/02/2022
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	25/02/2022	25/02/2022
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	25/02/2022	25/02/2022
					CT2750	Implantable cardiac electrical stimulators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	25/02/2022	25/02/2022
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	25/02/2022	25/02/2022
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	25/02/2022	25/02/2022
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	25/02/2022	25/02/2022
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	25/02/2022	25/02/2022
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	25/02/2022	25/02/2022
					CT2985	Active	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	25/02/2022	25/02/2022
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	25/02/2022	25/02/2022
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	25/02/2022	25/02/2022
					CT3001	Implantable metal	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	25/02/2022	25/02/2022
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	25/02/2022	25/02/2022
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	25/02/2022	25/02/2022
					CT3002	Implantable synthetic polymer	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	25/02/2022	25/02/2022
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	25/02/2022	25/02/2022
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	25/02/2022	25/02/2022
					CT3006	Implantable inorganic material	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	25/02/2022	25/02/2022
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	25/02/2022	25/02/2022
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	25/02/2022	25/02/2022
					CT314	Electrically powered	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	25/02/2022	25/02/2022
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	25/02/2022	25/02/2022
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	25/02/2022	25/02/2022
					CT321	Long-term surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	25/02/2022	25/02/2022
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	25/02/2022	25/02/2022
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	25/02/2022	25/02/2022
					CT449	Generators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	25/02/2022	25/02/2022
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	25/02/2022	25/02/2022
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	25/02/2022	25/02/2022
					CT981	Single use	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	25/02/2022	25/02/2022
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	25/02/2022	25/02/2022
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	25/02/2022	25/02/2022
					CT983	Surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	25/02/2022	25/02/2022

							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	25/02/2022	25/02/2022
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	25/02/2022	25/02/2022
6097	2022	All other reportable incidents	47267	Single-chamber implantable pacemaker, rate-responsive	CT141	Electrodes/Leads and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F27	Problem Identified During Non-Clinical Procedure	A070504	Premature Discharge of Battery		
					CT145	Cardiovascular devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F27	Problem Identified During Non-Clinical Procedure	A070504	Premature Discharge of Battery		
					CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F27	Problem Identified During Non-Clinical Procedure	A070504	Premature Discharge of Battery		
					CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F27	Problem Identified During Non-Clinical Procedure	A070504	Premature Discharge of Battery		
					CT233	Surgical	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F27	Problem Identified During Non-Clinical Procedure	A070504	Premature Discharge of Battery		
					CT2406	Implantable	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F27	Problem Identified During Non-Clinical Procedure	A070504	Premature Discharge of Battery		
					CT2407	Active implantable	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F27	Problem Identified During Non-Clinical Procedure	A070504	Premature Discharge of Battery		
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F27	Problem Identified During Non-Clinical Procedure	A070504	Premature Discharge of Battery		
					CT2727	Active implant devices/ancillaries	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F27	Problem Identified During Non-Clinical Procedure	A070504	Premature Discharge of Battery		
					CT2750	Implantable cardiac electrical stimulators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F27	Problem Identified During Non-Clinical Procedure	A070504	Premature Discharge of Battery		
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F27	Problem Identified During Non-Clinical Procedure	A070504	Premature Discharge of Battery		
					CT2985	Active	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F27	Problem Identified During Non-Clinical Procedure	A070504	Premature Discharge of Battery		
					CT3001	Implantable metal	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F27	Problem Identified During Non-Clinical Procedure	A070504	Premature Discharge of Battery		
					CT3002	Implantable synthetic polymer	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F27	Problem Identified During Non-Clinical Procedure	A070504	Premature Discharge of Battery		
					CT3006	Implantable inorganic material	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F27	Problem Identified During Non-Clinical Procedure	A070504	Premature Discharge of Battery		
					CT314	Electrically powered	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F27	Problem Identified During Non-Clinical Procedure	A070504	Premature Discharge of Battery		
					CT321	Long-term surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F27	Problem Identified During Non-Clinical Procedure	A070504	Premature Discharge of Battery		
					CT449	Generators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F27	Problem Identified During Non-Clinical Procedure	A070504	Premature Discharge of Battery		
					CT981	Single use	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F27	Problem Identified During Non-Clinical Procedure	A070504	Premature Discharge of Battery		
					CT983	Surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F27	Problem Identified During Non-Clinical Procedure	A070504	Premature Discharge of Battery		
6098	2021	All other reportable incidents	35852	Single-chamber implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E2104	Electric Shock	A071302	Inappropriate/Inadequate Shock/Stimulation		
							E2403	No Clinical Signs, Symptoms or Conditions	A071302	Inappropriate/Inadequate Shock/Stimulation		
							F26	No Health Consequences or Impact	A071302	Inappropriate/Inadequate Shock/Stimulation		
6099	2021	All other reportable incidents	47265	Dual-chamber implantable pacemaker, rate-responsive	CT141	Electrodes/Leads and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A072201	High Impedance		
							F08	Hospitalization or Prolonged Hospitalization	A072201	High Impedance		
							F1905	Device Revision or Replacement	A072201	High Impedance		
					CT145	Cardiovascular devices	E2403	No Clinical Signs, Symptoms or Conditions	A072201	High Impedance		
							F08	Hospitalization or Prolonged Hospitalization	A072201	High Impedance		
							F1905	Device Revision or Replacement	A072201	High Impedance		
					CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A072201	High Impedance		
							F08	Hospitalization or Prolonged Hospitalization	A072201	High Impedance		
							F1905	Device Revision or Replacement	A072201	High Impedance		
					CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A072201	High Impedance		
							F08	Hospitalization or Prolonged Hospitalization	A072201	High Impedance		
							F1905	Device Revision or Replacement	A072201	High Impedance		
					CT233	Surgical	E2403	No Clinical Signs, Symptoms or Conditions	A072201	High Impedance		
							F08	Hospitalization or Prolonged Hospitalization	A072201	High Impedance		
							F1905	Device Revision or Replacement	A072201	High Impedance		
					CT2406	Implantable	E2403	No Clinical Signs, Symptoms or Conditions	A072201	High Impedance		
							F08	Hospitalization or Prolonged Hospitalization	A072201	High Impedance		
							F1905	Device Revision or Replacement	A072201	High Impedance		
					CT2407	Active implantable	E2403	No Clinical Signs, Symptoms or Conditions	A072201	High Impedance		
							F08	Hospitalization or Prolonged Hospitalization	A072201	High Impedance		
							F1905	Device Revision or Replacement	A072201	High Impedance		
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A072201	High Impedance		
							F08	Hospitalization or Prolonged Hospitalization	A072201	High Impedance		
							F1905	Device Revision or Replacement	A072201	High Impedance		
					CT2727	Active implant devices/ancillaries	E2403	No Clinical Signs, Symptoms or Conditions	A072201	High Impedance		
							F08	Hospitalization or Prolonged Hospitalization	A072201	High Impedance		
							F1905	Device Revision or Replacement	A072201	High Impedance		
					CT2750	Implantable cardiac electrical stimulators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A072201	High Impedance		
							F08	Hospitalization or Prolonged Hospitalization	A072201	High Impedance		
							F1905	Device Revision or Replacement	A072201	High Impedance		
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A072201	High Impedance		
							F08	Hospitalization or Prolonged Hospitalization	A072201	High Impedance		
							F1905	Device Revision or Replacement	A072201	High Impedance		
					CT2985	Active	E2403	No Clinical Signs, Symptoms or Conditions	A072201	High Impedance		
							F08	Hospitalization or Prolonged Hospitalization	A072201	High Impedance		
							F1905	Device Revision or Replacement	A072201	High Impedance		
					CT3001	Implantable metal	E2403	No Clinical Signs, Symptoms or Conditions	A072201	High Impedance		
							F08	Hospitalization or Prolonged Hospitalization	A072201	High Impedance		
							F1905	Device Revision or Replacement	A072201	High Impedance		
					CT3002	Implantable synthetic polymer	E2403	No Clinical Signs, Symptoms or Conditions	A072201	High Impedance		
							F08	Hospitalization or Prolonged Hospitalization	A072201	High Impedance		
							F1905	Device Revision or Replacement	A072201	High Impedance		
					CT3006	Implantable inorganic material	E2403	No Clinical Signs, Symptoms or Conditions	A072201	High Impedance		

							F08	Hospitalization or Prolonged Hospitalization	A072201	High Impedance		
							F1905	Device Revision or Replacement	A072201	High Impedance		
					CT314	Electrically powered	E2403	No Clinical Signs, Symptoms or Conditions	A072201	High Impedance		
							F08	Hospitalization or Prolonged Hospitalization	A072201	High Impedance		
							F1905	Device Revision or Replacement	A072201	High Impedance		
					CT321	Long-term surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A072201	High Impedance		
							F08	Hospitalization or Prolonged Hospitalization	A072201	High Impedance		
							F1905	Device Revision or Replacement	A072201	High Impedance		
					CT449	Generators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A072201	High Impedance		
							F08	Hospitalization or Prolonged Hospitalization	A072201	High Impedance		
							F1905	Device Revision or Replacement	A072201	High Impedance		
					CT981	Single use	E2403	No Clinical Signs, Symptoms or Conditions	A072201	High Impedance		
							F08	Hospitalization or Prolonged Hospitalization	A072201	High Impedance		
							F1905	Device Revision or Replacement	A072201	High Impedance		
					CT983	Surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A072201	High Impedance		
							F08	Hospitalization or Prolonged Hospitalization	A072201	High Impedance		
							F1905	Device Revision or Replacement	A072201	High Impedance		
6100	2021	All other reportable incidents	35852	Single-chamber implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem	13/09/2021	13/09/2021
									A070504	Premature Discharge of Battery	13/09/2021	13/09/2021
							F08	Hospitalization or Prolonged Hospitalization	A0705	Battery Problem	13/09/2021	13/09/2021
									A070504	Premature Discharge of Battery	13/09/2021	13/09/2021
							F1905	Device Revision or Replacement	A0705	Battery Problem	13/09/2021	13/09/2021
									A070504	Premature Discharge of Battery	13/09/2021	13/09/2021
6101	2021	All other reportable incidents	36241	Pacemaker/defibrillator lead	CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070909	Oversensing		
									A072201	High Impedance		
									A090801	Signal Artifact/Noise		
							F15	Recognized Device or Procedural Complication	A070909	Oversensing		
									A072201	High Impedance		
									A090801	Signal Artifact/Noise		
6102	2021	All other reportable incidents	47265	Dual-chamber implantable pacemaker, rate-responsive	CT141	Electrodes/Leads and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070909	Oversensing		
									A072201	High Impedance		
							F15	Recognized Device or Procedural Complication	A070909	Oversensing		
									A072201	High Impedance		
					CT145	Cardiovascular devices	E2403	No Clinical Signs, Symptoms or Conditions	A070909	Oversensing		
									A072201	High Impedance		
							F15	Recognized Device or Procedural Complication	A070909	Oversensing		
									A072201	High Impedance		
					CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070909	Oversensing		
									A072201	High Impedance		
							F15	Recognized Device or Procedural Complication	A070909	Oversensing		
									A072201	High Impedance		
					CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A070909	Oversensing		
									A072201	High Impedance		
							F15	Recognized Device or Procedural Complication	A070909	Oversensing		
									A072201	High Impedance		
					CT233	Surgical	E2403	No Clinical Signs, Symptoms or Conditions	A070909	Oversensing		
									A072201	High Impedance		
							F15	Recognized Device or Procedural Complication	A070909	Oversensing		
									A072201	High Impedance		
					CT2406	Implantable	E2403	No Clinical Signs, Symptoms or Conditions	A070909	Oversensing		
									A072201	High Impedance		
							F15	Recognized Device or Procedural Complication	A070909	Oversensing		
									A072201	High Impedance		
					CT2407	Active implantable	E2403	No Clinical Signs, Symptoms or Conditions	A070909	Oversensing		
									A072201	High Impedance		
							F15	Recognized Device or Procedural Complication	A070909	Oversensing		
									A072201	High Impedance		
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A070909	Oversensing		
									A072201	High Impedance		
							F15	Recognized Device or Procedural Complication	A070909	Oversensing		
									A072201	High Impedance		
					CT2727	Active implant devices/ancillaries	E2403	No Clinical Signs, Symptoms or Conditions	A070909	Oversensing		
									A072201	High Impedance		
							F15	Recognized Device or Procedural Complication	A070909	Oversensing		
									A072201	High Impedance		
					CT2750	Implantable cardiac electrical stimulators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070909	Oversensing		
									A072201	High Impedance		
							F15	Recognized Device or Procedural Complication	A070909	Oversensing		
									A072201	High Impedance		
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A070909	Oversensing		
									A072201	High Impedance		
							F15	Recognized Device or Procedural Complication	A070909	Oversensing		
									A072201	High Impedance		
					CT2985	Active	E2403	No Clinical Signs, Symptoms or Conditions	A070909	Oversensing		
									A072201	High Impedance		
							F15	Recognized Device or Procedural Complication	A070909	Oversensing		
									A072201	High Impedance		
					CT3001	Implantable metal	E2403	No Clinical Signs, Symptoms or Conditions	A070909	Oversensing		
									A072201	High Impedance		
							F15	Recognized Device or Procedural Complication	A070909	Oversensing		
									A072201	High Impedance		
					CT3002	Implantable synthetic polymer	E2403	No Clinical Signs, Symptoms or Conditions	A070909	Oversensing		
									A072201	High Impedance		
							F15	Recognized Device or Procedural Complication	A070909	Oversensing		
									A072201	High Impedance		
					CT3006	Implantable inorganic material	E2403	No Clinical Signs, Symptoms or Conditions	A070909	Oversensing		
									A072201	High Impedance		
							F15	Recognized Device or Procedural Complication	A070909	Oversensing		

								A072201	High Impedance			
				CT314	Electrically powered	E2403	No Clinical Signs, Symptoms or Conditions	A070909	Oversensing			
								A072201	High Impedance			
						F15	Recognized Device or Procedural Complication	A070909	Oversensing			
								A072201	High Impedance			
				CT321	Long-term surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A070909	Oversensing			
								A072201	High Impedance			
						F15	Recognized Device or Procedural Complication	A070909	Oversensing			
								A072201	High Impedance			
				CT449	Generators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070909	Oversensing			
								A072201	High Impedance			
						F15	Recognized Device or Procedural Complication	A070909	Oversensing			
								A072201	High Impedance			
				CT981	Single use	E2403	No Clinical Signs, Symptoms or Conditions	A070909	Oversensing			
								A072201	High Impedance			
						F15	Recognized Device or Procedural Complication	A070909	Oversensing			
								A072201	High Impedance			
				CT983	Surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A070909	Oversensing			
								A072201	High Impedance			
						F15	Recognized Device or Procedural Complication	A070909	Oversensing			
								A072201	High Impedance			
6103	2022	All other reportable incidents	37265	Dual-chamber implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A090805	Incorrect Measurement		
							F26	No Health Consequences or Impact	A090805	Incorrect Measurement		
6104	2021	All other reportable incidents	47265	Dual-chamber implantable pacemaker, rate-responsive	CT141	Electrodes/Leads and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A072201	High Impedance		
							F08	Hospitalization or Prolonged Hospitalization	A072201	High Impedance		
							F1905	Device Revision or Replacement	A072201	High Impedance		
					CT145	Cardiovascular devices	E2403	No Clinical Signs, Symptoms or Conditions	A072201	High Impedance		
							F08	Hospitalization or Prolonged Hospitalization	A072201	High Impedance		
							F1905	Device Revision or Replacement	A072201	High Impedance		
					CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A072201	High Impedance		
							F08	Hospitalization or Prolonged Hospitalization	A072201	High Impedance		
							F1905	Device Revision or Replacement	A072201	High Impedance		
					CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A072201	High Impedance		
							F08	Hospitalization or Prolonged Hospitalization	A072201	High Impedance		
							F1905	Device Revision or Replacement	A072201	High Impedance		
					CT233	Surgical	E2403	No Clinical Signs, Symptoms or Conditions	A072201	High Impedance		
							F08	Hospitalization or Prolonged Hospitalization	A072201	High Impedance		
							F1905	Device Revision or Replacement	A072201	High Impedance		
					CT2406	Implantable	E2403	No Clinical Signs, Symptoms or Conditions	A072201	High Impedance		
							F08	Hospitalization or Prolonged Hospitalization	A072201	High Impedance		
							F1905	Device Revision or Replacement	A072201	High Impedance		
					CT2407	Active implantable	E2403	No Clinical Signs, Symptoms or Conditions	A072201	High Impedance		
							F08	Hospitalization or Prolonged Hospitalization	A072201	High Impedance		
							F1905	Device Revision or Replacement	A072201	High Impedance		
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A072201	High Impedance		
							F08	Hospitalization or Prolonged Hospitalization	A072201	High Impedance		
							F1905	Device Revision or Replacement	A072201	High Impedance		
					CT2727	Active implant devices/ancillaries	E2403	No Clinical Signs, Symptoms or Conditions	A072201	High Impedance		
							F08	Hospitalization or Prolonged Hospitalization	A072201	High Impedance		
							F1905	Device Revision or Replacement	A072201	High Impedance		
					CT2750	Implantable cardiac electrical stimulators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A072201	High Impedance		
							F08	Hospitalization or Prolonged Hospitalization	A072201	High Impedance		
							F1905	Device Revision or Replacement	A072201	High Impedance		
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A072201	High Impedance		
							F08	Hospitalization or Prolonged Hospitalization	A072201	High Impedance		
							F1905	Device Revision or Replacement	A072201	High Impedance		
					CT2985	Active	E2403	No Clinical Signs, Symptoms or Conditions	A072201	High Impedance		
							F08	Hospitalization or Prolonged Hospitalization	A072201	High Impedance		
							F1905	Device Revision or Replacement	A072201	High Impedance		
					CT3001	Implantable metal	E2403	No Clinical Signs, Symptoms or Conditions	A072201	High Impedance		
							F08	Hospitalization or Prolonged Hospitalization	A072201	High Impedance		
							F1905	Device Revision or Replacement	A072201	High Impedance		
					CT3002	Implantable synthetic polymer	E2403	No Clinical Signs, Symptoms or Conditions	A072201	High Impedance		
							F08	Hospitalization or Prolonged Hospitalization	A072201	High Impedance		
							F1905	Device Revision or Replacement	A072201	High Impedance		
					CT3006	Implantable inorganic material	E2403	No Clinical Signs, Symptoms or Conditions	A072201	High Impedance		
							F08	Hospitalization or Prolonged Hospitalization	A072201	High Impedance		
							F1905	Device Revision or Replacement	A072201	High Impedance		
					CT314	Electrically powered	E2403	No Clinical Signs, Symptoms or Conditions	A072201	High Impedance		
							F08	Hospitalization or Prolonged Hospitalization	A072201	High Impedance		
							F1905	Device Revision or Replacement	A072201	High Impedance		
					CT321	Long-term surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A072201	High Impedance		
							F08	Hospitalization or Prolonged Hospitalization	A072201	High Impedance		
							F1905	Device Revision or Replacement	A072201	High Impedance		
					CT449	Generators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A072201	High Impedance		
							F08	Hospitalization or Prolonged Hospitalization	A072201	High Impedance		
							F1905	Device Revision or Replacement	A072201	High Impedance		
					CT981	Single use	E2403	No Clinical Signs, Symptoms or Conditions	A072201	High Impedance		
							F08	Hospitalization or Prolonged Hospitalization	A072201	High Impedance		
							F1905	Device Revision or Replacement	A072201	High Impedance		
					CT983	Surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A072201	High Impedance		
							F08	Hospitalization or Prolonged Hospitalization	A072201	High Impedance		
							F1905	Device Revision or Replacement	A072201	High Impedance		
6105	2021	All other reportable incidents	35852	Single-chamber implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	19/09/2021	19/09/2021
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	19/09/2021	19/09/2021
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	19/09/2021	19/09/2021
6106	2022	All other reportable incidents	36241	Pacemaker/defibrillator lead	CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A072201	High Impedance		
							F2203	Imaging Required	A072201	High Impedance		

6107	2022	All other reportable incidents	47265	Dual-chamber implantable pacemaker, rate-responsive	CT141	Electrodes/Leads and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	14/04/2022	14/04/2022
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	14/04/2022	14/04/2022
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	14/04/2022	14/04/2022
					CT145	Cardiovascular devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	14/04/2022	14/04/2022
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	14/04/2022	14/04/2022
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	14/04/2022	14/04/2022
					CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	14/04/2022	14/04/2022
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	14/04/2022	14/04/2022
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	14/04/2022	14/04/2022
					CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	14/04/2022	14/04/2022
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	14/04/2022	14/04/2022
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	14/04/2022	14/04/2022
					CT233	Surgical	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	14/04/2022	14/04/2022
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	14/04/2022	14/04/2022
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	14/04/2022	14/04/2022
					CT2406	Implantable	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	14/04/2022	14/04/2022
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	14/04/2022	14/04/2022
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	14/04/2022	14/04/2022
					CT2407	Active implantable	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	14/04/2022	14/04/2022
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	14/04/2022	14/04/2022
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	14/04/2022	14/04/2022
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	14/04/2022	14/04/2022
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	14/04/2022	14/04/2022
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	14/04/2022	14/04/2022
					CT2727	Active implant devices/ancillaries	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	14/04/2022	14/04/2022
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	14/04/2022	14/04/2022
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	14/04/2022	14/04/2022
					CT2750	Implantable cardiac electrical stimulators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	14/04/2022	14/04/2022
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	14/04/2022	14/04/2022
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	14/04/2022	14/04/2022
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	14/04/2022	14/04/2022
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	14/04/2022	14/04/2022
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	14/04/2022	14/04/2022
					CT2985	Active	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	14/04/2022	14/04/2022
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	14/04/2022	14/04/2022
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	14/04/2022	14/04/2022
					CT3001	Implantable metal	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	14/04/2022	14/04/2022
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	14/04/2022	14/04/2022
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	14/04/2022	14/04/2022
					CT3002	Implantable synthetic polymer	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	14/04/2022	14/04/2022
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	14/04/2022	14/04/2022
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	14/04/2022	14/04/2022
					CT3006	Implantable inorganic material	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	14/04/2022	14/04/2022
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	14/04/2022	14/04/2022
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	14/04/2022	14/04/2022
					CT314	Electrically powered	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	14/04/2022	14/04/2022
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	14/04/2022	14/04/2022
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	14/04/2022	14/04/2022
					CT321	Long-term surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	14/04/2022	14/04/2022
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	14/04/2022	14/04/2022
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	14/04/2022	14/04/2022
					CT449	Generators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	14/04/2022	14/04/2022
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	14/04/2022	14/04/2022
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	14/04/2022	14/04/2022
					CT981	Single use	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	14/04/2022	14/04/2022
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	14/04/2022	14/04/2022
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	14/04/2022	14/04/2022
					CT983	Surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	14/04/2022	14/04/2022
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	14/04/2022	14/04/2022
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	14/04/2022	14/04/2022
6108	2022	All other reportable incidents	35223	Endocardial/interventricular septal pacing lead	CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A072001	Electromagnetic Interference		
									A090801	Signal Artifact/Noise		
							F26	No Health Consequences or Impact	A072001	Electromagnetic Interference		
									A090801	Signal Artifact/Noise		
6109	2022	All other reportable incidents	35223	Endocardial/interventricular septal pacing lead	CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A07	Electrical /Electronic Property Problem		
									A072202	Low Impedance		
							F08	Hospitalization or Prolonged Hospitalization	A07	Electrical /Electronic Property Problem		
									A072202	Low Impedance		
							F12	Serious Injury/ Illness/ Impairment	A07	Electrical /Electronic Property Problem		
									A072202	Low Impedance		
							F1905	Device Revision or Replacement	A07	Electrical /Electronic Property Problem		
									A072202	Low Impedance		
							F23	Unexpected Medical Intervention	A07	Electrical /Electronic Property Problem		
									A072202	Low Impedance		
6110	2022	All other reportable incidents	47265	Dual-chamber implantable pacemaker, rate-responsive	CT141	Electrodes/Leads and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery		
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery		
					CT145	Cardiovascular devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery		
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery		
					CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery		
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery		
					CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery		
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery		
					CT233	Surgical	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery		

							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery		
					CT2406	Implantable	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery		
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery		
					CT2407	Active implantable	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery		
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery		
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery		
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery		
					CT2727	Active implant devices/ancillaries	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery		
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery		
					CT2750	Implantable cardiac electrical stimulators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery		
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery		
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery		
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery		
					CT2985	Active	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery		
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery		
					CT3001	Implantable metal	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery		
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery		
					CT3002	Implantable synthetic polymer	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery		
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery		
					CT3006	Implantable inorganic material	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery		
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery		
					CT314	Electrically powered	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery		
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery		
					CT321	Long-term surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery		
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery		
					CT449	Generators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery		
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery		
					CT981	Single use	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery		
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery		
					CT983	Surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery		
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery		
6111	2022	All other reportable incidents	47267	Single-chamber implantable pacemaker, rate-responsive	CT141	Electrodes/Leads and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	21/02/2022	21/02/2022
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	21/02/2022	21/02/2022
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	21/02/2022	21/02/2022
					CT145	Cardiovascular devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	21/02/2022	21/02/2022
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	21/02/2022	21/02/2022
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	21/02/2022	21/02/2022
					CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	21/02/2022	21/02/2022
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	21/02/2022	21/02/2022
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	21/02/2022	21/02/2022
					CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	21/02/2022	21/02/2022
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	21/02/2022	21/02/2022
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	21/02/2022	21/02/2022
					CT233	Surgical	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	21/02/2022	21/02/2022
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	21/02/2022	21/02/2022
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	21/02/2022	21/02/2022
					CT2406	Implantable	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	21/02/2022	21/02/2022
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	21/02/2022	21/02/2022
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	21/02/2022	21/02/2022
					CT2407	Active implantable	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	21/02/2022	21/02/2022
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	21/02/2022	21/02/2022
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	21/02/2022	21/02/2022
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	21/02/2022	21/02/2022
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	21/02/2022	21/02/2022
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	21/02/2022	21/02/2022
					CT2727	Active implant devices/ancillaries	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	21/02/2022	21/02/2022
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	21/02/2022	21/02/2022
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	21/02/2022	21/02/2022
					CT2750	Implantable cardiac electrical stimulators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	21/02/2022	21/02/2022
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	21/02/2022	21/02/2022
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	21/02/2022	21/02/2022
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	21/02/2022	21/02/2022
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	21/02/2022	21/02/2022
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	21/02/2022	21/02/2022
					CT2985	Active	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	21/02/2022	21/02/2022
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	21/02/2022	21/02/2022
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	21/02/2022	21/02/2022
					CT3001	Implantable metal	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	21/02/2022	21/02/2022
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	21/02/2022	21/02/2022
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	21/02/2022	21/02/2022
					CT3002	Implantable synthetic polymer	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	21/02/2022	21/02/2022
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	21/02/2022	21/02/2022
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	21/02/2022	21/02/2022

					CT3006	Implantable inorganic material	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	21/02/2022	21/02/2022
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	21/02/2022	21/02/2022
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	21/02/2022	21/02/2022
					CT314	Electrically powered	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	21/02/2022	21/02/2022
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	21/02/2022	21/02/2022
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	21/02/2022	21/02/2022
					CT321	Long-term surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	21/02/2022	21/02/2022
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	21/02/2022	21/02/2022
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	21/02/2022	21/02/2022
					CT449	Generators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	21/02/2022	21/02/2022
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	21/02/2022	21/02/2022
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	21/02/2022	21/02/2022
					CT981	Single use	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	21/02/2022	21/02/2022
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	21/02/2022	21/02/2022
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	21/02/2022	21/02/2022
					CT983	Surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	21/02/2022	21/02/2022
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	21/02/2022	21/02/2022
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	21/02/2022	21/02/2022
6112	2022	All other reportable incidents	36241	Pacemaker/defibrillator lead	CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A072201	High Impedance		
							F26	No Health Consequences or Impact	A072201	High Impedance		
6113	2022	All other reportable incidents	36241	Pacemaker/defibrillator lead	CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A072201	High Impedance		
							F26	No Health Consequences or Impact	A072201	High Impedance		
6114	2022	All other reportable incidents	35852	Single-chamber implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A010402	Migration		
									A072201	High Impedance		
							F26	No Health Consequences or Impact	A010402	Migration		
									A072201	High Impedance		
6115	2021	All other reportable incidents	47265	Dual-chamber implantable pacemaker, rate-responsive	CT141	Electrodes/Leads and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A072201	High Impedance		
							F15	Recognized Device or Procedural Complication	A072201	High Impedance		
					CT145	Cardiovascular devices	E2403	No Clinical Signs, Symptoms or Conditions	A072201	High Impedance		
							F15	Recognized Device or Procedural Complication	A072201	High Impedance		
					CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A072201	High Impedance		
							F15	Recognized Device or Procedural Complication	A072201	High Impedance		
					CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A072201	High Impedance		
							F15	Recognized Device or Procedural Complication	A072201	High Impedance		
					CT233	Surgical	E2403	No Clinical Signs, Symptoms or Conditions	A072201	High Impedance		
							F15	Recognized Device or Procedural Complication	A072201	High Impedance		
					CT2406	Implantable	E2403	No Clinical Signs, Symptoms or Conditions	A072201	High Impedance		
							F15	Recognized Device or Procedural Complication	A072201	High Impedance		
					CT2407	Active implantable	E2403	No Clinical Signs, Symptoms or Conditions	A072201	High Impedance		
							F15	Recognized Device or Procedural Complication	A072201	High Impedance		
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A072201	High Impedance		
							F15	Recognized Device or Procedural Complication	A072201	High Impedance		
					CT2727	Active implant devices/ancillaries	E2403	No Clinical Signs, Symptoms or Conditions	A072201	High Impedance		
							F15	Recognized Device or Procedural Complication	A072201	High Impedance		
					CT2750	Implantable cardiac electrical stimulators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A072201	High Impedance		
							F15	Recognized Device or Procedural Complication	A072201	High Impedance		
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A072201	High Impedance		
							F15	Recognized Device or Procedural Complication	A072201	High Impedance		
					CT2985	Active	E2403	No Clinical Signs, Symptoms or Conditions	A072201	High Impedance		
							F15	Recognized Device or Procedural Complication	A072201	High Impedance		
					CT3001	Implantable metal	E2403	No Clinical Signs, Symptoms or Conditions	A072201	High Impedance		
							F15	Recognized Device or Procedural Complication	A072201	High Impedance		
					CT3002	Implantable synthetic polymer	E2403	No Clinical Signs, Symptoms or Conditions	A072201	High Impedance		
							F15	Recognized Device or Procedural Complication	A072201	High Impedance		
					CT3006	Implantable inorganic material	E2403	No Clinical Signs, Symptoms or Conditions	A072201	High Impedance		
							F15	Recognized Device or Procedural Complication	A072201	High Impedance		
					CT314	Electrically powered	E2403	No Clinical Signs, Symptoms or Conditions	A072201	High Impedance		
							F15	Recognized Device or Procedural Complication	A072201	High Impedance		
					CT321	Long-term surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A072201	High Impedance		
							F15	Recognized Device or Procedural Complication	A072201	High Impedance		
					CT449	Generators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A072201	High Impedance		
							F15	Recognized Device or Procedural Complication	A072201	High Impedance		
					CT981	Single use	E2403	No Clinical Signs, Symptoms or Conditions	A072201	High Impedance		
							F15	Recognized Device or Procedural Complication	A072201	High Impedance		
					CT983	Surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A072201	High Impedance		
							F15	Recognized Device or Procedural Complication	A072201	High Impedance		
6116	2021	All other reportable incidents	35852	Single-chamber implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A070909	Oversensing		
							F26	No Health Consequences or Impact	A070909	Oversensing		
6117	2021	All other reportable incidents	37265	Dual-chamber implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A072201	High Impedance		
							F26	No Health Consequences or Impact	A072201	High Impedance		
6118	2021	All other reportable incidents	47263	Cardiac resynchronization therapy implantable pacemaker	CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
6119	2021	All other reportable incidents	47265	Dual-chamber implantable pacemaker, rate-responsive	CT141	Electrodes/Leads and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT145	Cardiovascular devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT233	Surgical	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		

					CT2406	Implantable	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT2407	Active implantable	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT2727	Active implant devices/ancillaries	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT2750	Implantable cardiac electrical stimulators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT2985	Active	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT3001	Implantable metal	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT3002	Implantable synthetic polymer	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT3006	Implantable inorganic material	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT314	Electrically powered	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT321	Long-term surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT449	Generators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT981	Single use	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT983	Surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
6120	2021	All other reportable incidents	47265	Dual-chamber implantable pacemaker, rate-responsive	CT141	Electrodes/Leads and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT145	Cardiovascular devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT233	Surgical	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT2406	Implantable	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT2407	Active implantable	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT2727	Active implant devices/ancillaries	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT2750	Implantable cardiac electrical stimulators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT2985	Active	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT3001	Implantable metal	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT3002	Implantable synthetic polymer	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT3006	Implantable inorganic material	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT314	Electrically powered	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT321	Long-term surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT449	Generators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT981	Single use	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT983	Surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
6121	2021	All other reportable incidents	37265	Dual-chamber implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	23/08/2021	23/08/2021
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	23/08/2021	23/08/2021
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	23/08/2021	23/08/2021
6122	2021	All other reportable incidents	47265	Dual-chamber implantable pacemaker, rate-responsive	CT141	Electrodes/Leads and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	28/09/2021	28/09/2021
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	28/09/2021	28/09/2021
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	28/09/2021	28/09/2021
					CT145	Cardiovascular devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	28/09/2021	28/09/2021
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	28/09/2021	28/09/2021
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	28/09/2021	28/09/2021
					CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	28/09/2021	28/09/2021
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	28/09/2021	28/09/2021
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	28/09/2021	28/09/2021
					CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	28/09/2021	28/09/2021
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	28/09/2021	28/09/2021
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	28/09/2021	28/09/2021
					CT233	Surgical	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	28/09/2021	28/09/2021
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	28/09/2021	28/09/2021
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	28/09/2021	28/09/2021

					CT2406	Implantable	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	28/09/2021	28/09/2021
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	28/09/2021	28/09/2021
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	28/09/2021	28/09/2021
					CT2407	Active implantable	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	28/09/2021	28/09/2021
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	28/09/2021	28/09/2021
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	28/09/2021	28/09/2021
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	28/09/2021	28/09/2021
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	28/09/2021	28/09/2021
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	28/09/2021	28/09/2021
					CT2727	Active implant devices/ancillaries	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	28/09/2021	28/09/2021
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	28/09/2021	28/09/2021
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	28/09/2021	28/09/2021
					CT2750	Implantable cardiac electrical stimulators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	28/09/2021	28/09/2021
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	28/09/2021	28/09/2021
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	28/09/2021	28/09/2021
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	28/09/2021	28/09/2021
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	28/09/2021	28/09/2021
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	28/09/2021	28/09/2021
					CT2985	Active	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	28/09/2021	28/09/2021
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	28/09/2021	28/09/2021
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	28/09/2021	28/09/2021
					CT3001	Implantable metal	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	28/09/2021	28/09/2021
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	28/09/2021	28/09/2021
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	28/09/2021	28/09/2021
					CT3002	Implantable synthetic polymer	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	28/09/2021	28/09/2021
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	28/09/2021	28/09/2021
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	28/09/2021	28/09/2021
					CT3006	Implantable inorganic material	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	28/09/2021	28/09/2021
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	28/09/2021	28/09/2021
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	28/09/2021	28/09/2021
					CT314	Electrically powered	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	28/09/2021	28/09/2021
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	28/09/2021	28/09/2021
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	28/09/2021	28/09/2021
					CT321	Long-term surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	28/09/2021	28/09/2021
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	28/09/2021	28/09/2021
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	28/09/2021	28/09/2021
					CT449	Generators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	28/09/2021	28/09/2021
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	28/09/2021	28/09/2021
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	28/09/2021	28/09/2021
					CT981	Single use	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	28/09/2021	28/09/2021
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	28/09/2021	28/09/2021
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	28/09/2021	28/09/2021
					CT983	Surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	28/09/2021	28/09/2021
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	28/09/2021	28/09/2021
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	28/09/2021	28/09/2021
6123	2022	All other reportable incidents	17882	Physiologic-monitoring defibrillation system	CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A0801	Failure to Calibrate		
							F27	Problem Identified During Non-Clinical Procedure	A0801	Failure to Calibrate		
6124	2022	All other reportable incidents	47270	Cardiac resynchronization therapy implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A070909	Oversensing		
									A0712	Pacing Problem		
							F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A070909	Oversensing		
									A0712	Pacing Problem		
							F1001	Absence of Treatment	A070909	Oversensing		
									A0712	Pacing Problem		
6125	2022	All other reportable incidents	47263	Cardiac resynchronization therapy implantable pacemaker	CT1684	Cardiac pacemakers and associated devices	E0138	Undesired Nerve Stimulation	A0908	Incorrect, Inadequate or Imprecise Result or Readings		
							E2403	No Clinical Signs, Symptoms or Conditions	A0908	Incorrect, Inadequate or Imprecise Result or Readings		
							F26	No Health Consequences or Impact	A0908	Incorrect, Inadequate or Imprecise Result or Readings		
					CT1685	Cardiac pacemakers	E0138	Undesired Nerve Stimulation	A0908	Incorrect, Inadequate or Imprecise Result or Readings		
							E2403	No Clinical Signs, Symptoms or Conditions	A0908	Incorrect, Inadequate or Imprecise Result or Readings		
							F26	No Health Consequences or Impact	A0908	Incorrect, Inadequate or Imprecise Result or Readings		
					CT257	Implantable cardiac pacemakers	E0138	Undesired Nerve Stimulation	A0908	Incorrect, Inadequate or Imprecise Result or Readings		
							E2403	No Clinical Signs, Symptoms or Conditions	A0908	Incorrect, Inadequate or Imprecise Result or Readings		
							F26	No Health Consequences or Impact	A0908	Incorrect, Inadequate or Imprecise Result or Readings		
					CT2866	Implantable cardiac electrical stimulators	E0138	Undesired Nerve Stimulation	A0908	Incorrect, Inadequate or Imprecise Result or Readings		
							E2403	No Clinical Signs, Symptoms or Conditions	A0908	Incorrect, Inadequate or Imprecise Result or Readings		
							F26	No Health Consequences or Impact	A0908	Incorrect, Inadequate or Imprecise Result or Readings		
6126	2022	All other reportable incidents	47263	Cardiac resynchronization therapy implantable pacemaker	CT1684	Cardiac pacemakers and associated devices	E0138	Undesired Nerve Stimulation	A0908	Incorrect, Inadequate or Imprecise Result or Readings		
							E2403	No Clinical Signs, Symptoms or Conditions	A0908	Incorrect, Inadequate or Imprecise Result or Readings		
							F26	No Health Consequences or Impact	A0908	Incorrect, Inadequate or Imprecise Result or Readings		
					CT1685	Cardiac pacemakers	E0138	Undesired Nerve Stimulation	A0908	Incorrect, Inadequate or Imprecise Result or Readings		
							E2403	No Clinical Signs, Symptoms or Conditions	A0908	Incorrect, Inadequate or Imprecise Result or Readings		
							F26	No Health Consequences or Impact	A0908	Incorrect, Inadequate or Imprecise Result or Readings		
					CT257	Implantable cardiac pacemakers	E0138	Undesired Nerve Stimulation	A0908	Incorrect, Inadequate or Imprecise Result or Readings		

						E2403	No Clinical Signs, Symptoms or Conditions	A0908	Incorrect, Inadequate or Imprecise Result or Readings			
						F26	No Health Consequences or Impact	A0908	Incorrect, Inadequate or Imprecise Result or Readings			
					CT2866	Implantable cardiac electrical stimulators	E0138	Undesired Nerve Stimulation	A0908	Incorrect, Inadequate or Imprecise Result or Readings		
							E2403	No Clinical Signs, Symptoms or Conditions	A0908	Incorrect, Inadequate or Imprecise Result or Readings		
							F26	No Health Consequences or Impact	A0908	Incorrect, Inadequate or Imprecise Result or Readings		
6127	2022	All other reportable incidents	47263	Cardiac resynchronization therapy implantable pacemaker	CT1684	Cardiac pacemakers and associated devices	E0138	Undesired Nerve Stimulation	A0908	Incorrect, Inadequate or Imprecise Result or Readings		
							E2403	No Clinical Signs, Symptoms or Conditions	A0908	Incorrect, Inadequate or Imprecise Result or Readings		
							F08	Hospitalization or Prolonged Hospitalization	A0908	Incorrect, Inadequate or Imprecise Result or Readings		
							F1905	Device Revision or Replacement	A0908	Incorrect, Inadequate or Imprecise Result or Readings		
					CT1685	Cardiac pacemakers	E0138	Undesired Nerve Stimulation	A0908	Incorrect, Inadequate or Imprecise Result or Readings		
							E2403	No Clinical Signs, Symptoms or Conditions	A0908	Incorrect, Inadequate or Imprecise Result or Readings		
							F08	Hospitalization or Prolonged Hospitalization	A0908	Incorrect, Inadequate or Imprecise Result or Readings		
							F1905	Device Revision or Replacement	A0908	Incorrect, Inadequate or Imprecise Result or Readings		
					CT257	Implantable cardiac pacemakers	E0138	Undesired Nerve Stimulation	A0908	Incorrect, Inadequate or Imprecise Result or Readings		
							E2403	No Clinical Signs, Symptoms or Conditions	A0908	Incorrect, Inadequate or Imprecise Result or Readings		
							F08	Hospitalization or Prolonged Hospitalization	A0908	Incorrect, Inadequate or Imprecise Result or Readings		
							F1905	Device Revision or Replacement	A0908	Incorrect, Inadequate or Imprecise Result or Readings		
					CT2866	Implantable cardiac electrical stimulators	E0138	Undesired Nerve Stimulation	A0908	Incorrect, Inadequate or Imprecise Result or Readings		
							E2403	No Clinical Signs, Symptoms or Conditions	A0908	Incorrect, Inadequate or Imprecise Result or Readings		
							F08	Hospitalization or Prolonged Hospitalization	A0908	Incorrect, Inadequate or Imprecise Result or Readings		
							F1905	Device Revision or Replacement	A0908	Incorrect, Inadequate or Imprecise Result or Readings		
6128	2022	All other reportable incidents	47270	Cardiac resynchronization therapy implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A070101	Failure to Capture		
									A070909	Oversensing		
							F26	No Health Consequences or Impact	A070101	Failure to Capture		
									A070909	Oversensing		
6129	2021	All other reportable incidents	35223	Endocardial/interventricular septal pacing lead	CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A040101	Fracture		
									A070909	Oversensing		
									A0712	Pacing Problem		
									A090801	Signal Artifact/Noise		
							F26	No Health Consequences or Impact	A040101	Fracture		
									A070909	Oversensing		
									A0712	Pacing Problem		
									A090801	Signal Artifact/Noise		
6130	2021	All other reportable incidents	60910	Coronary venous pacing lead	CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A0701	Capturing Problem		
									A070101	Failure to Capture		
									A0722	Impedance Problem		
							F26	No Health Consequences or Impact	A0701	Capturing Problem		
									A070101	Failure to Capture		
									A0722	Impedance Problem		
6131	2021	All other reportable incidents	47265	Dual-chamber implantable pacemaker, rate-responsive	CT141	Electrodes/Leads and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	20/08/2021	20/08/2021
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	20/08/2021	20/08/2021
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	20/08/2021	20/08/2021
					CT145	Cardiovascular devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	20/08/2021	20/08/2021
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	20/08/2021	20/08/2021
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	20/08/2021	20/08/2021
					CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	20/08/2021	20/08/2021
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	20/08/2021	20/08/2021
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	20/08/2021	20/08/2021
					CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	20/08/2021	20/08/2021
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	20/08/2021	20/08/2021
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	20/08/2021	20/08/2021
					CT233	Surgical	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	20/08/2021	20/08/2021
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	20/08/2021	20/08/2021
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	20/08/2021	20/08/2021
					CT2406	Implantable	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	20/08/2021	20/08/2021
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	20/08/2021	20/08/2021
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	20/08/2021	20/08/2021
					CT2407	Active implantable	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	20/08/2021	20/08/2021
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	20/08/2021	20/08/2021
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	20/08/2021	20/08/2021
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	20/08/2021	20/08/2021
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	20/08/2021	20/08/2021
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	20/08/2021	20/08/2021
					CT2727	Active implant devices/ancillaries	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	20/08/2021	20/08/2021
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	20/08/2021	20/08/2021
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	20/08/2021	20/08/2021
					CT2750	Implantable cardiac electrical stimulators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	20/08/2021	20/08/2021
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	20/08/2021	20/08/2021
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	20/08/2021	20/08/2021
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	20/08/2021	20/08/2021
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	20/08/2021	20/08/2021
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	20/08/2021	20/08/2021

					CT2985	Active	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	20/08/2021	20/08/2021
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	20/08/2021	20/08/2021
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	20/08/2021	20/08/2021
					CT3001	Implantable metal	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	20/08/2021	20/08/2021
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	20/08/2021	20/08/2021
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	20/08/2021	20/08/2021
					CT3002	Implantable synthetic polymer	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	20/08/2021	20/08/2021
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	20/08/2021	20/08/2021
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	20/08/2021	20/08/2021
					CT3006	Implantable inorganic material	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	20/08/2021	20/08/2021
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	20/08/2021	20/08/2021
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	20/08/2021	20/08/2021
					CT314	Electrically powered	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	20/08/2021	20/08/2021
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	20/08/2021	20/08/2021
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	20/08/2021	20/08/2021
					CT321	Long-term surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	20/08/2021	20/08/2021
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	20/08/2021	20/08/2021
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	20/08/2021	20/08/2021
					CT449	Generators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	20/08/2021	20/08/2021
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	20/08/2021	20/08/2021
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	20/08/2021	20/08/2021
					CT981	Single use	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	20/08/2021	20/08/2021
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	20/08/2021	20/08/2021
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	20/08/2021	20/08/2021
					CT983	Surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	20/08/2021	20/08/2021
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	20/08/2021	20/08/2021
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	20/08/2021	20/08/2021
6132	2021	All other reportable incidents	35852	Single-chamber implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A070909	Oversensing		
							F26	No Health Consequences or Impact	A070909	Oversensing		
6133	2022	All other reportable incidents	47265	Dual-chamber implantable pacemaker, rate-responsive	CT141	Electrodes/Leads and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT145	Cardiovascular devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT233	Surgical	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT2406	Implantable	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT2407	Active implantable	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT2727	Active implant devices/ancillaries	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT2750	Implantable cardiac electrical stimulators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT2985	Active	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT3001	Implantable metal	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT3002	Implantable synthetic polymer	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT3006	Implantable inorganic material	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT314	Electrically powered	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT321	Long-term surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT449	Generators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT981	Single use	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT983	Surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
6134	2022	All other reportable incidents	47265	Dual-chamber implantable pacemaker, rate-responsive	CT141	Electrodes/Leads and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	25/05/2022	25/05/2022
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	25/05/2022	25/05/2022
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	25/05/2022	25/05/2022
					CT145	Cardiovascular devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	25/05/2022	25/05/2022
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	25/05/2022	25/05/2022
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	25/05/2022	25/05/2022
					CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	25/05/2022	25/05/2022
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	25/05/2022	25/05/2022
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	25/05/2022	25/05/2022
					CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	25/05/2022	25/05/2022
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	25/05/2022	25/05/2022
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	25/05/2022	25/05/2022
					CT233	Surgical	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	25/05/2022	25/05/2022
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	25/05/2022	25/05/2022
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	25/05/2022	25/05/2022
					CT2406	Implantable	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	25/05/2022	25/05/2022
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	25/05/2022	25/05/2022
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	25/05/2022	25/05/2022
					CT2407	Active implantable	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	25/05/2022	25/05/2022

							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	25/05/2022	25/05/2022
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	25/05/2022	25/05/2022
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	25/05/2022	25/05/2022
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	25/05/2022	25/05/2022
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	25/05/2022	25/05/2022
					CT2727	Active implant devices/ancillaries	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	25/05/2022	25/05/2022
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	25/05/2022	25/05/2022
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	25/05/2022	25/05/2022
					CT2750	Implantable cardiac electrical stimulators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	25/05/2022	25/05/2022
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	25/05/2022	25/05/2022
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	25/05/2022	25/05/2022
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	25/05/2022	25/05/2022
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	25/05/2022	25/05/2022
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	25/05/2022	25/05/2022
					CT2985	Active	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	25/05/2022	25/05/2022
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	25/05/2022	25/05/2022
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	25/05/2022	25/05/2022
					CT3001	Implantable metal	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	25/05/2022	25/05/2022
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	25/05/2022	25/05/2022
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	25/05/2022	25/05/2022
					CT3002	Implantable synthetic polymer	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	25/05/2022	25/05/2022
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	25/05/2022	25/05/2022
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	25/05/2022	25/05/2022
					CT3006	Implantable inorganic material	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	25/05/2022	25/05/2022
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	25/05/2022	25/05/2022
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	25/05/2022	25/05/2022
					CT314	Electrically powered	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	25/05/2022	25/05/2022
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	25/05/2022	25/05/2022
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	25/05/2022	25/05/2022
					CT321	Long-term surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	25/05/2022	25/05/2022
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	25/05/2022	25/05/2022
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	25/05/2022	25/05/2022
					CT449	Generators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	25/05/2022	25/05/2022
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	25/05/2022	25/05/2022
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	25/05/2022	25/05/2022
					CT981	Single use	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	25/05/2022	25/05/2022
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	25/05/2022	25/05/2022
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	25/05/2022	25/05/2022
					CT983	Surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	25/05/2022	25/05/2022
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	25/05/2022	25/05/2022
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	25/05/2022	25/05/2022
6135	2022	All other reportable incidents	17882	Physiologic-monitoring defibrillation system	CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070908	Failure to Sense		
							F27	Problem Identified During Non-Clinical Procedure	A070908	Failure to Sense		
6136	2022	All other reportable incidents	47267	Single-chamber implantable pacemaker, rate-responsive	CT141	Electrodes/Leads and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A0908	Incorrect, Inadequate or Imprecise Result or Readings	03/03/2022	03/03/2022
							F08	Hospitalization or Prolonged Hospitalization	A0908	Incorrect, Inadequate or Imprecise Result or Readings	03/03/2022	03/03/2022
							F1905	Device Revision or Replacement	A0908	Incorrect, Inadequate or Imprecise Result or Readings	03/03/2022	03/03/2022
					CT145	Cardiovascular devices	E2403	No Clinical Signs, Symptoms or Conditions	A0908	Incorrect, Inadequate or Imprecise Result or Readings	03/03/2022	03/03/2022
							F08	Hospitalization or Prolonged Hospitalization	A0908	Incorrect, Inadequate or Imprecise Result or Readings	03/03/2022	03/03/2022
							F1905	Device Revision or Replacement	A0908	Incorrect, Inadequate or Imprecise Result or Readings	03/03/2022	03/03/2022
					CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A0908	Incorrect, Inadequate or Imprecise Result or Readings	03/03/2022	03/03/2022
							F08	Hospitalization or Prolonged Hospitalization	A0908	Incorrect, Inadequate or Imprecise Result or Readings	03/03/2022	03/03/2022
							F1905	Device Revision or Replacement	A0908	Incorrect, Inadequate or Imprecise Result or Readings	03/03/2022	03/03/2022
					CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A0908	Incorrect, Inadequate or Imprecise Result or Readings	03/03/2022	03/03/2022
							F08	Hospitalization or Prolonged Hospitalization	A0908	Incorrect, Inadequate or Imprecise Result or Readings	03/03/2022	03/03/2022
							F1905	Device Revision or Replacement	A0908	Incorrect, Inadequate or Imprecise Result or Readings	03/03/2022	03/03/2022
					CT233	Surgical	E2403	No Clinical Signs, Symptoms or Conditions	A0908	Incorrect, Inadequate or Imprecise Result or Readings	03/03/2022	03/03/2022
							F08	Hospitalization or Prolonged Hospitalization	A0908	Incorrect, Inadequate or Imprecise Result or Readings	03/03/2022	03/03/2022
							F1905	Device Revision or Replacement	A0908	Incorrect, Inadequate or Imprecise Result or Readings	03/03/2022	03/03/2022
					CT2406	Implantable	E2403	No Clinical Signs, Symptoms or Conditions	A0908	Incorrect, Inadequate or Imprecise Result or Readings	03/03/2022	03/03/2022
							F08	Hospitalization or Prolonged Hospitalization	A0908	Incorrect, Inadequate or Imprecise Result or Readings	03/03/2022	03/03/2022
							F1905	Device Revision or Replacement	A0908	Incorrect, Inadequate or Imprecise Result or Readings	03/03/2022	03/03/2022
					CT2407	Active implantable	E2403	No Clinical Signs, Symptoms or Conditions	A0908	Incorrect, Inadequate or Imprecise Result or Readings	03/03/2022	03/03/2022
							F08	Hospitalization or Prolonged Hospitalization	A0908	Incorrect, Inadequate or Imprecise Result or Readings	03/03/2022	03/03/2022
							F1905	Device Revision or Replacement	A0908	Incorrect, Inadequate or Imprecise Result or Readings	03/03/2022	03/03/2022
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A0908	Incorrect, Inadequate or Imprecise Result or Readings	03/03/2022	03/03/2022
							F08	Hospitalization or Prolonged Hospitalization	A0908	Incorrect, Inadequate or Imprecise Result or Readings	03/03/2022	03/03/2022
							F1905	Device Revision or Replacement	A0908	Incorrect, Inadequate or Imprecise Result or Readings	03/03/2022	03/03/2022
					CT2727	Active implant devices/ancillaries	E2403	No Clinical Signs, Symptoms or Conditions	A0908	Incorrect, Inadequate or Imprecise Result or Readings	03/03/2022	03/03/2022

							F08	Hospitalization or Prolonged Hospitalization	A0908	Incorrect, Inadequate or Imprecise Result or Readings	03/03/2022	03/03/2022
							F1905	Device Revision or Replacement	A0908	Incorrect, Inadequate or Imprecise Result or Readings	03/03/2022	03/03/2022
					CT2750	Implantable cardiac electrical stimulators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A0908	Incorrect, Inadequate or Imprecise Result or Readings	03/03/2022	03/03/2022
							F08	Hospitalization or Prolonged Hospitalization	A0908	Incorrect, Inadequate or Imprecise Result or Readings	03/03/2022	03/03/2022
							F1905	Device Revision or Replacement	A0908	Incorrect, Inadequate or Imprecise Result or Readings	03/03/2022	03/03/2022
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A0908	Incorrect, Inadequate or Imprecise Result or Readings	03/03/2022	03/03/2022
							F08	Hospitalization or Prolonged Hospitalization	A0908	Incorrect, Inadequate or Imprecise Result or Readings	03/03/2022	03/03/2022
							F1905	Device Revision or Replacement	A0908	Incorrect, Inadequate or Imprecise Result or Readings	03/03/2022	03/03/2022
					CT2985	Active	E2403	No Clinical Signs, Symptoms or Conditions	A0908	Incorrect, Inadequate or Imprecise Result or Readings	03/03/2022	03/03/2022
							F08	Hospitalization or Prolonged Hospitalization	A0908	Incorrect, Inadequate or Imprecise Result or Readings	03/03/2022	03/03/2022
							F1905	Device Revision or Replacement	A0908	Incorrect, Inadequate or Imprecise Result or Readings	03/03/2022	03/03/2022
					CT3001	Implantable metal	E2403	No Clinical Signs, Symptoms or Conditions	A0908	Incorrect, Inadequate or Imprecise Result or Readings	03/03/2022	03/03/2022
							F08	Hospitalization or Prolonged Hospitalization	A0908	Incorrect, Inadequate or Imprecise Result or Readings	03/03/2022	03/03/2022
							F1905	Device Revision or Replacement	A0908	Incorrect, Inadequate or Imprecise Result or Readings	03/03/2022	03/03/2022
					CT3002	Implantable synthetic polymer	E2403	No Clinical Signs, Symptoms or Conditions	A0908	Incorrect, Inadequate or Imprecise Result or Readings	03/03/2022	03/03/2022
							F08	Hospitalization or Prolonged Hospitalization	A0908	Incorrect, Inadequate or Imprecise Result or Readings	03/03/2022	03/03/2022
							F1905	Device Revision or Replacement	A0908	Incorrect, Inadequate or Imprecise Result or Readings	03/03/2022	03/03/2022
					CT3006	Implantable inorganic material	E2403	No Clinical Signs, Symptoms or Conditions	A0908	Incorrect, Inadequate or Imprecise Result or Readings	03/03/2022	03/03/2022
							F08	Hospitalization or Prolonged Hospitalization	A0908	Incorrect, Inadequate or Imprecise Result or Readings	03/03/2022	03/03/2022
							F1905	Device Revision or Replacement	A0908	Incorrect, Inadequate or Imprecise Result or Readings	03/03/2022	03/03/2022
					CT314	Electrically powered	E2403	No Clinical Signs, Symptoms or Conditions	A0908	Incorrect, Inadequate or Imprecise Result or Readings	03/03/2022	03/03/2022
							F08	Hospitalization or Prolonged Hospitalization	A0908	Incorrect, Inadequate or Imprecise Result or Readings	03/03/2022	03/03/2022
							F1905	Device Revision or Replacement	A0908	Incorrect, Inadequate or Imprecise Result or Readings	03/03/2022	03/03/2022
					CT321	Long-term surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A0908	Incorrect, Inadequate or Imprecise Result or Readings	03/03/2022	03/03/2022
							F08	Hospitalization or Prolonged Hospitalization	A0908	Incorrect, Inadequate or Imprecise Result or Readings	03/03/2022	03/03/2022
							F1905	Device Revision or Replacement	A0908	Incorrect, Inadequate or Imprecise Result or Readings	03/03/2022	03/03/2022
					CT449	Generators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A0908	Incorrect, Inadequate or Imprecise Result or Readings	03/03/2022	03/03/2022
							F08	Hospitalization or Prolonged Hospitalization	A0908	Incorrect, Inadequate or Imprecise Result or Readings	03/03/2022	03/03/2022
							F1905	Device Revision or Replacement	A0908	Incorrect, Inadequate or Imprecise Result or Readings	03/03/2022	03/03/2022
					CT981	Single use	E2403	No Clinical Signs, Symptoms or Conditions	A0908	Incorrect, Inadequate or Imprecise Result or Readings	03/03/2022	03/03/2022
							F08	Hospitalization or Prolonged Hospitalization	A0908	Incorrect, Inadequate or Imprecise Result or Readings	03/03/2022	03/03/2022
							F1905	Device Revision or Replacement	A0908	Incorrect, Inadequate or Imprecise Result or Readings	03/03/2022	03/03/2022
					CT983	Surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A0908	Incorrect, Inadequate or Imprecise Result or Readings	03/03/2022	03/03/2022
							F08	Hospitalization or Prolonged Hospitalization	A0908	Incorrect, Inadequate or Imprecise Result or Readings	03/03/2022	03/03/2022
							F1905	Device Revision or Replacement	A0908	Incorrect, Inadequate or Imprecise Result or Readings	03/03/2022	03/03/2022
6137	2022	All other reportable incidents	47265	Dual-chamber implantable pacemaker, rate-responsive	CT141	Electrodes/Leads and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070909	Oversensing		
							F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A070909	Oversensing		
					CT145	Cardiovascular devices	E2403	No Clinical Signs, Symptoms or Conditions	A070909	Oversensing		
							F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A070909	Oversensing		
					CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070909	Oversensing		
							F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A070909	Oversensing		
					CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A070909	Oversensing		
							F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A070909	Oversensing		
					CT233	Surgical	E2403	No Clinical Signs, Symptoms or Conditions	A070909	Oversensing		
							F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A070909	Oversensing		
					CT2406	Implantable	E2403	No Clinical Signs, Symptoms or Conditions	A070909	Oversensing		
							F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A070909	Oversensing		
					CT2407	Active implantable	E2403	No Clinical Signs, Symptoms or Conditions	A070909	Oversensing		
							F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A070909	Oversensing		
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A070909	Oversensing		
							F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A070909	Oversensing		
					CT2727	Active implant devices/ancillaries	E2403	No Clinical Signs, Symptoms or Conditions	A070909	Oversensing		
							F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A070909	Oversensing		
					CT2750	Implantable cardiac electrical stimulators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070909	Oversensing		
							F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A070909	Oversensing		
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A070909	Oversensing		
							F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A070909	Oversensing		
					CT2985	Active	E2403	No Clinical Signs, Symptoms or Conditions	A070909	Oversensing		
							F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A070909	Oversensing		
					CT3001	Implantable metal	E2403	No Clinical Signs, Symptoms or Conditions	A070909	Oversensing		
							F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A070909	Oversensing		

					CT3002	Implantable synthetic polymer	E2403	No Clinical Signs, Symptoms or Conditions	A070909	Oversensing		
							F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A070909	Oversensing		
					CT3006	Implantable inorganic material	E2403	No Clinical Signs, Symptoms or Conditions	A070909	Oversensing		
							F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A070909	Oversensing		
					CT314	Electrically powered	E2403	No Clinical Signs, Symptoms or Conditions	A070909	Oversensing		
							F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A070909	Oversensing		
					CT321	Long-term surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A070909	Oversensing		
							F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A070909	Oversensing		
					CT449	Generators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070909	Oversensing		
							F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A070909	Oversensing		
					CT981	Single use	E2403	No Clinical Signs, Symptoms or Conditions	A070909	Oversensing		
							F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A070909	Oversensing		
					CT983	Surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A070909	Oversensing		
							F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A070909	Oversensing		
6138	2022	All other reportable incidents	35852	Single-chamber implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E2104	Electric Shock	A010402	Migration		
									A070909	Oversensing		
									A071302	Inappropriate/Inadequate Shock/Stimulation		
							E2403	No Clinical Signs, Symptoms or Conditions	A010402	Migration		
									A070909	Oversensing		
									A071302	Inappropriate/Inadequate Shock/Stimulation		
							F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A010402	Migration		
									A070909	Oversensing		
									A071302	Inappropriate/Inadequate Shock/Stimulation		
							F2203	Imaging Required	A010402	Migration		
									A070909	Oversensing		
									A071302	Inappropriate/Inadequate Shock/Stimulation		
6139	2021	All other reportable incidents	36241	Pacemaker/defibrillator lead	CT1684	Cardiac pacemakers and associated devices	E0604	Cardiac Perforation	A24	No Apparent Device Problem		
							E0619	Pericardial Effusion	A24	No Apparent Device Problem		
							E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem		
							F2203	Imaging Required	A24	No Apparent Device Problem		
							F28	Appropriate Health Impact Term/Code not Available	A24	No Apparent Device Problem		
6140	2021	All other reportable incidents	36241	Pacemaker/defibrillator lead	CT1684	Cardiac pacemakers and associated devices	E0604	Cardiac Perforation	A24	No Apparent Device Problem		
							E0619	Pericardial Effusion	A24	No Apparent Device Problem		
							E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem		
							F2203	Imaging Required	A24	No Apparent Device Problem		
							F28	Appropriate Health Impact Term/Code not Available	A24	No Apparent Device Problem		
6141	2021	All other reportable incidents	17882	Physiologic-monitoring defibrillation system	CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A090812	Missing Test Results		
							F27	Problem Identified During Non-Clinical Procedure	A090812	Missing Test Results		
6142	2021	All other reportable incidents	17882	Physiologic-monitoring defibrillation system	CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A090801	Signal Artifact/Noise		
							F27	Problem Identified During Non-Clinical Procedure	A090801	Signal Artifact/Noise		
6143	2021	All other reportable incidents	17882	Physiologic-monitoring defibrillation system	CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070803	Failure to Power Up		
							F27	Problem Identified During Non-Clinical Procedure	A070803	Failure to Power Up		
6144	2022	All other reportable incidents	35852	Single-chamber implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A072201	High Impedance		
							F26	No Health Consequences or Impact	A072201	High Impedance		
6145	2022	All other reportable incidents	36241	Pacemaker/defibrillator lead	CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A072201	High Impedance	24/02/2022	24/02/2022
							F1908	Prolonged Surgery	A072201	High Impedance	24/02/2022	24/02/2022
6146	2021	All other reportable incidents	47265	Dual-chamber implantable pacemaker, rate-responsive	CT141	Electrodes/Leads and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
					CT145	Cardiovascular devices	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
					CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
					CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
					CT233	Surgical	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
					CT2406	Implantable	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
					CT2407	Active implantable	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
					CT2727	Active implant devices/ancillaries	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
					CT2750	Implantable cardiac electrical stimulators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		

							F26	No Health Consequences or Impact	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
					CT2985	Active	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
					CT3001	Implantable metal	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
					CT3002	Implantable synthetic polymer	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
					CT3006	Implantable inorganic material	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
					CT314	Electrically powered	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
					CT321	Long-term surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
					CT449	Generators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
					CT981	Single use	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
					CT983	Surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
6147	2021	All other reportable incidents	35852	Single-chamber implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem	09/08/2021	09/08/2021
									A070504	Premature Discharge of Battery	09/08/2021	09/08/2021
							F08	Hospitalization or Prolonged Hospitalization	A0705	Battery Problem	09/08/2021	09/08/2021
									A070504	Premature Discharge of Battery	09/08/2021	09/08/2021
							F1905	Device Revision or Replacement	A0705	Battery Problem	09/08/2021	09/08/2021
									A070504	Premature Discharge of Battery	09/08/2021	09/08/2021
6148	2022	All other reportable incidents	35039	Epicardial pacing lead	CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070101	Failure to Capture		
									A070102	High Capture Threshold		
									A070909	Oversensing		
									A090801	Signal Artifact/Noise		
							F26	No Health Consequences or Impact	A070101	Failure to Capture		
									A070102	High Capture Threshold		
									A070909	Oversensing		
									A090801	Signal Artifact/Noise		
6149	2021	All other reportable incidents	35039	Epicardial pacing lead	CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070909	Oversensing	16/07/2021	16/07/2021
									A090801	Signal Artifact/Noise	16/07/2021	16/07/2021
							F26	No Health Consequences or Impact	A070909	Oversensing	16/07/2021	16/07/2021
									A090801	Signal Artifact/Noise	16/07/2021	16/07/2021
6150	2021	All other reportable incidents	35852	Single-chamber implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
									A0908	Incorrect, Inadequate or Imprecise Result or Readings		
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery		
									A0908	Incorrect, Inadequate or Imprecise Result or Readings		
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery		
									A0908	Incorrect, Inadequate or Imprecise Result or Readings		
6151	2022	All other reportable incidents	35223	Endocardial/interventricular septal pacing lead	CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A040101	Fracture		
									A070909	Oversensing		
									A072201	High Impedance		
									A090801	Signal Artifact/Noise		
							F26	No Health Consequences or Impact	A040101	Fracture		
									A070909	Oversensing		
									A072201	High Impedance		
									A090801	Signal Artifact/Noise		
6152	2021	All other reportable incidents	35852	Single-chamber implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	02/03/2021	02/03/2021
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	02/03/2021	02/03/2021
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	02/03/2021	02/03/2021
6153	2021	Death	17882	Physiologic-monitoring defibrillation system	CT1684	Cardiac pacemakers and associated devices	E2401	Insufficient Information	A071301	Failure to Deliver Shock/Stimulation		
							F05	Delay to Treatment/ Therapy	A071301	Failure to Deliver Shock/Stimulation		
6154	2021	All other reportable incidents	37265	Dual-chamber implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F1901	Additional Surgery	A070504	Premature Discharge of Battery		
6155	2022	All other reportable incidents	47263	Cardiac resynchronization therapy implantable pacemaker	CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	21/06/2022	21/06/2022
							F1901	Additional Surgery	A070504	Premature Discharge of Battery	21/06/2022	21/06/2022
					CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	21/06/2022	21/06/2022
							F1901	Additional Surgery	A070504	Premature Discharge of Battery	21/06/2022	21/06/2022
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	21/06/2022	21/06/2022
							F1901	Additional Surgery	A070504	Premature Discharge of Battery	21/06/2022	21/06/2022
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	21/06/2022	21/06/2022
							F1901	Additional Surgery	A070504	Premature Discharge of Battery	21/06/2022	21/06/2022

6156	2021	All other reportable incidents	17882	Physiologic-monitoring defibrillation system	CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070803	Failure to Power Up		
							F27	Problem Identified During Non-Clinical Procedure	A070803	Failure to Power Up		
6157	2021	All other reportable incidents	47106	Drug-eluting endocardial pacing lead	CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070101	Failure to Capture		
									A070102	High Capture Threshold		
							F01	Change in Therapeutic Response	A070101	Failure to Capture		
									A070102	High Capture Threshold		
6158	2022	All other reportable incidents	47265	Dual-chamber implantable pacemaker, rate-responsive	CT141	Electrodes/Leads and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A090801	Signal Artifact/Noise		
							F26	No Health Consequences or Impact	A090801	Signal Artifact/Noise		
					CT145	Cardiovascular devices	E2403	No Clinical Signs, Symptoms or Conditions	A090801	Signal Artifact/Noise		
							F26	No Health Consequences or Impact	A090801	Signal Artifact/Noise		
					CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A090801	Signal Artifact/Noise		
							F26	No Health Consequences or Impact	A090801	Signal Artifact/Noise		
					CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A090801	Signal Artifact/Noise		
							F26	No Health Consequences or Impact	A090801	Signal Artifact/Noise		
					CT233	Surgical	E2403	No Clinical Signs, Symptoms or Conditions	A090801	Signal Artifact/Noise		
							F26	No Health Consequences or Impact	A090801	Signal Artifact/Noise		
					CT2406	Implantable	E2403	No Clinical Signs, Symptoms or Conditions	A090801	Signal Artifact/Noise		
							F26	No Health Consequences or Impact	A090801	Signal Artifact/Noise		
					CT2407	Active implantable	E2403	No Clinical Signs, Symptoms or Conditions	A090801	Signal Artifact/Noise		
							F26	No Health Consequences or Impact	A090801	Signal Artifact/Noise		
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A090801	Signal Artifact/Noise		
							F26	No Health Consequences or Impact	A090801	Signal Artifact/Noise		
					CT2727	Active implant devices/ancillaries	E2403	No Clinical Signs, Symptoms or Conditions	A090801	Signal Artifact/Noise		
							F26	No Health Consequences or Impact	A090801	Signal Artifact/Noise		
					CT2750	Implantable cardiac electrical stimulators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A090801	Signal Artifact/Noise		
							F26	No Health Consequences or Impact	A090801	Signal Artifact/Noise		
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A090801	Signal Artifact/Noise		
							F26	No Health Consequences or Impact	A090801	Signal Artifact/Noise		
					CT2985	Active	E2403	No Clinical Signs, Symptoms or Conditions	A090801	Signal Artifact/Noise		
							F26	No Health Consequences or Impact	A090801	Signal Artifact/Noise		
					CT3001	Implantable metal	E2403	No Clinical Signs, Symptoms or Conditions	A090801	Signal Artifact/Noise		
							F26	No Health Consequences or Impact	A090801	Signal Artifact/Noise		
					CT3002	Implantable synthetic polymer	E2403	No Clinical Signs, Symptoms or Conditions	A090801	Signal Artifact/Noise		
							F26	No Health Consequences or Impact	A090801	Signal Artifact/Noise		
					CT3006	Implantable inorganic material	E2403	No Clinical Signs, Symptoms or Conditions	A090801	Signal Artifact/Noise		
							F26	No Health Consequences or Impact	A090801	Signal Artifact/Noise		
					CT314	Electrically powered	E2403	No Clinical Signs, Symptoms or Conditions	A090801	Signal Artifact/Noise		
							F26	No Health Consequences or Impact	A090801	Signal Artifact/Noise		
					CT321	Long-term surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A090801	Signal Artifact/Noise		
							F26	No Health Consequences or Impact	A090801	Signal Artifact/Noise		
					CT449	Generators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A090801	Signal Artifact/Noise		
							F26	No Health Consequences or Impact	A090801	Signal Artifact/Noise		
					CT981	Single use	E2403	No Clinical Signs, Symptoms or Conditions	A090801	Signal Artifact/Noise		
							F26	No Health Consequences or Impact	A090801	Signal Artifact/Noise		
					CT983	Surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A090801	Signal Artifact/Noise		
							F26	No Health Consequences or Impact	A090801	Signal Artifact/Noise		
6159	2022	All other reportable incidents	47270	Cardiac resynchronization therapy implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A070909	Oversensing		
							F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A070909	Oversensing		
6160	2022	All other reportable incidents	47265	Dual-chamber implantable pacemaker, rate-responsive	CT141	Electrodes/Leads and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070909	Oversensing		
							F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A070909	Oversensing		
					CT145	Cardiovascular devices	E2403	No Clinical Signs, Symptoms or Conditions	A070909	Oversensing		
							F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A070909	Oversensing		
					CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070909	Oversensing		
							F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A070909	Oversensing		
					CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A070909	Oversensing		
							F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A070909	Oversensing		
					CT233	Surgical	E2403	No Clinical Signs, Symptoms or Conditions	A070909	Oversensing		
							F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A070909	Oversensing		
					CT2406	Implantable	E2403	No Clinical Signs, Symptoms or Conditions	A070909	Oversensing		
							F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A070909	Oversensing		
					CT2407	Active implantable	E2403	No Clinical Signs, Symptoms or Conditions	A070909	Oversensing		
							F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A070909	Oversensing		
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A070909	Oversensing		
							F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A070909	Oversensing		
					CT2727	Active implant devices/ancillaries	E2403	No Clinical Signs, Symptoms or Conditions	A070909	Oversensing		
							F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A070909	Oversensing		
					CT2750	Implantable cardiac electrical stimulators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070909	Oversensing		
							F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A070909	Oversensing		
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A070909	Oversensing		
							F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A070909	Oversensing		
					CT2985	Active	E2403	No Clinical Signs, Symptoms or Conditions	A070909	Oversensing		
							F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A070909	Oversensing		
					CT3001	Implantable metal	E2403	No Clinical Signs, Symptoms or Conditions	A070909	Oversensing		
							F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A070909	Oversensing		
					CT3002	Implantable synthetic polymer	E2403	No Clinical Signs, Symptoms or Conditions	A070909	Oversensing		
							F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A070909	Oversensing		
					CT3006	Implantable inorganic material	E2403	No Clinical Signs, Symptoms or Conditions	A070909	Oversensing		
							F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A070909	Oversensing		
					CT314	Electrically powered	E2403	No Clinical Signs, Symptoms or Conditions	A070909	Oversensing		
							F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A070909	Oversensing		
					CT321	Long-term surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A070909	Oversensing		
							F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A070909	Oversensing		
					CT449	Generators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070909	Oversensing		
							F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A070909	Oversensing		
					CT981	Single use	E2403	No Clinical Signs, Symptoms or Conditions	A070909	Oversensing		
							F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A070909	Oversensing		
					CT983	Surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A070909	Oversensing		
							F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A070909	Oversensing		

6161	2022	All other reportable incidents	47265	Dual-chamber implantable pacemaker, rate-responsive	CT141	Electrodes/Leads and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem	13/05/2022	13/05/2022
									A070504	Premature Discharge of Battery	13/05/2022	13/05/2022
							F08	Hospitalization or Prolonged Hospitalization	A0705	Battery Problem	13/05/2022	13/05/2022
									A070504	Premature Discharge of Battery	13/05/2022	13/05/2022
							F1905	Device Revision or Replacement	A0705	Battery Problem	13/05/2022	13/05/2022
									A070504	Premature Discharge of Battery	13/05/2022	13/05/2022
					CT145	Cardiovascular devices	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem	13/05/2022	13/05/2022
									A070504	Premature Discharge of Battery	13/05/2022	13/05/2022
							F08	Hospitalization or Prolonged Hospitalization	A0705	Battery Problem	13/05/2022	13/05/2022
									A070504	Premature Discharge of Battery	13/05/2022	13/05/2022
							F1905	Device Revision or Replacement	A0705	Battery Problem	13/05/2022	13/05/2022
									A070504	Premature Discharge of Battery	13/05/2022	13/05/2022
					CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem	13/05/2022	13/05/2022
									A070504	Premature Discharge of Battery	13/05/2022	13/05/2022
							F08	Hospitalization or Prolonged Hospitalization	A0705	Battery Problem	13/05/2022	13/05/2022
									A070504	Premature Discharge of Battery	13/05/2022	13/05/2022
							F1905	Device Revision or Replacement	A0705	Battery Problem	13/05/2022	13/05/2022
									A070504	Premature Discharge of Battery	13/05/2022	13/05/2022
					CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem	13/05/2022	13/05/2022
									A070504	Premature Discharge of Battery	13/05/2022	13/05/2022
							F08	Hospitalization or Prolonged Hospitalization	A0705	Battery Problem	13/05/2022	13/05/2022
									A070504	Premature Discharge of Battery	13/05/2022	13/05/2022
							F1905	Device Revision or Replacement	A0705	Battery Problem	13/05/2022	13/05/2022
									A070504	Premature Discharge of Battery	13/05/2022	13/05/2022
					CT233	Surgical	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem	13/05/2022	13/05/2022
									A070504	Premature Discharge of Battery	13/05/2022	13/05/2022
							F08	Hospitalization or Prolonged Hospitalization	A0705	Battery Problem	13/05/2022	13/05/2022
									A070504	Premature Discharge of Battery	13/05/2022	13/05/2022
							F1905	Device Revision or Replacement	A0705	Battery Problem	13/05/2022	13/05/2022
									A070504	Premature Discharge of Battery	13/05/2022	13/05/2022
					CT2406	Implantable	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem	13/05/2022	13/05/2022
									A070504	Premature Discharge of Battery	13/05/2022	13/05/2022
							F08	Hospitalization or Prolonged Hospitalization	A0705	Battery Problem	13/05/2022	13/05/2022
									A070504	Premature Discharge of Battery	13/05/2022	13/05/2022
							F1905	Device Revision or Replacement	A0705	Battery Problem	13/05/2022	13/05/2022
									A070504	Premature Discharge of Battery	13/05/2022	13/05/2022
					CT2407	Active implantable	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem	13/05/2022	13/05/2022
									A070504	Premature Discharge of Battery	13/05/2022	13/05/2022
							F08	Hospitalization or Prolonged Hospitalization	A0705	Battery Problem	13/05/2022	13/05/2022
									A070504	Premature Discharge of Battery	13/05/2022	13/05/2022
							F1905	Device Revision or Replacement	A0705	Battery Problem	13/05/2022	13/05/2022
									A070504	Premature Discharge of Battery	13/05/2022	13/05/2022
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem	13/05/2022	13/05/2022
									A070504	Premature Discharge of Battery	13/05/2022	13/05/2022
							F08	Hospitalization or Prolonged Hospitalization	A0705	Battery Problem	13/05/2022	13/05/2022
									A070504	Premature Discharge of Battery	13/05/2022	13/05/2022
							F1905	Device Revision or Replacement	A0705	Battery Problem	13/05/2022	13/05/2022
									A070504	Premature Discharge of Battery	13/05/2022	13/05/2022
					CT2727	Active implant devices/ancillaries	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem	13/05/2022	13/05/2022
									A070504	Premature Discharge of Battery	13/05/2022	13/05/2022
							F08	Hospitalization or Prolonged Hospitalization	A0705	Battery Problem	13/05/2022	13/05/2022
									A070504	Premature Discharge of Battery	13/05/2022	13/05/2022
							F1905	Device Revision or Replacement	A0705	Battery Problem	13/05/2022	13/05/2022
									A070504	Premature Discharge of Battery	13/05/2022	13/05/2022
					CT2750	Implantable cardiac electrical stimulators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem	13/05/2022	13/05/2022
									A070504	Premature Discharge of Battery	13/05/2022	13/05/2022
							F08	Hospitalization or Prolonged Hospitalization	A0705	Battery Problem	13/05/2022	13/05/2022
									A070504	Premature Discharge of Battery	13/05/2022	13/05/2022
							F1905	Device Revision or Replacement	A0705	Battery Problem	13/05/2022	13/05/2022
									A070504	Premature Discharge of Battery	13/05/2022	13/05/2022
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem	13/05/2022	13/05/2022
									A070504	Premature Discharge of Battery	13/05/2022	13/05/2022
							F08	Hospitalization or Prolonged Hospitalization	A0705	Battery Problem	13/05/2022	13/05/2022
									A070504	Premature Discharge of Battery	13/05/2022	13/05/2022
							F1905	Device Revision or Replacement	A0705	Battery Problem	13/05/2022	13/05/2022
									A070504	Premature Discharge of Battery	13/05/2022	13/05/2022
					CT2985	Active	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem	13/05/2022	13/05/2022
									A070504	Premature Discharge of Battery	13/05/2022	13/05/2022
							F08	Hospitalization or Prolonged Hospitalization	A0705	Battery Problem	13/05/2022	13/05/2022
									A070504	Premature Discharge of Battery	13/05/2022	13/05/2022
							F1905	Device Revision or Replacement	A0705	Battery Problem	13/05/2022	13/05/2022
									A070504	Premature Discharge of Battery	13/05/2022	13/05/2022
					CT3001	Implantable metal	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem	13/05/2022	13/05/2022
									A070504	Premature Discharge of Battery	13/05/2022	13/05/2022
							F08	Hospitalization or Prolonged Hospitalization	A0705	Battery Problem	13/05/2022	13/05/2022
									A070504	Premature Discharge of Battery	13/05/2022	13/05/2022
							F1905	Device Revision or Replacement	A0705	Battery Problem	13/05/2022	13/05/2022
									A070504	Premature Discharge of Battery	13/05/2022	13/05/2022
					CT3002	Implantable synthetic polymer	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem	13/05/2022	13/05/2022
									A070504	Premature Discharge of Battery	13/05/2022	13/05/2022
							F08	Hospitalization or Prolonged Hospitalization	A0705	Battery Problem	13/05/2022	13/05/2022
									A070504	Premature Discharge of Battery	13/05/2022	13/05/2022
							F1905	Device Revision or Replacement	A0705	Battery Problem	13/05/2022	13/05/2022
									A070504	Premature Discharge of Battery	13/05/2022	13/05/2022
					CT3006	Implantable inorganic material	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem	13/05/2022	13/05/2022
									A070504	Premature Discharge of Battery	13/05/2022	13/05/2022
							F08	Hospitalization or Prolonged Hospitalization	A0705	Battery Problem	13/05/2022	13/05/2022
									A070504	Premature Discharge of Battery	13/05/2022	13/05/2022

						F1905	Device Revision or Replacement	A0705	Battery Problem	13/05/2022	13/05/2022
								A070504	Premature Discharge of Battery	13/05/2022	13/05/2022
					CT314	Electrically powered	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem	13/05/2022	13/05/2022
									A070504	Premature Discharge of Battery	13/05/2022	13/05/2022
							F08	Hospitalization or Prolonged Hospitalization	A0705	Battery Problem	13/05/2022	13/05/2022
									A070504	Premature Discharge of Battery	13/05/2022	13/05/2022
							F1905	Device Revision or Replacement	A0705	Battery Problem	13/05/2022	13/05/2022
									A070504	Premature Discharge of Battery	13/05/2022	13/05/2022
					CT321	Long-term surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem	13/05/2022	13/05/2022
									A070504	Premature Discharge of Battery	13/05/2022	13/05/2022
							F08	Hospitalization or Prolonged Hospitalization	A0705	Battery Problem	13/05/2022	13/05/2022
									A070504	Premature Discharge of Battery	13/05/2022	13/05/2022
							F1905	Device Revision or Replacement	A0705	Battery Problem	13/05/2022	13/05/2022
									A070504	Premature Discharge of Battery	13/05/2022	13/05/2022
					CT449	Generators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem	13/05/2022	13/05/2022
									A070504	Premature Discharge of Battery	13/05/2022	13/05/2022
							F08	Hospitalization or Prolonged Hospitalization	A0705	Battery Problem	13/05/2022	13/05/2022
									A070504	Premature Discharge of Battery	13/05/2022	13/05/2022
							F1905	Device Revision or Replacement	A0705	Battery Problem	13/05/2022	13/05/2022
									A070504	Premature Discharge of Battery	13/05/2022	13/05/2022
					CT981	Single use	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem	13/05/2022	13/05/2022
									A070504	Premature Discharge of Battery	13/05/2022	13/05/2022
							F08	Hospitalization or Prolonged Hospitalization	A0705	Battery Problem	13/05/2022	13/05/2022
									A070504	Premature Discharge of Battery	13/05/2022	13/05/2022
							F1905	Device Revision or Replacement	A0705	Battery Problem	13/05/2022	13/05/2022
									A070504	Premature Discharge of Battery	13/05/2022	13/05/2022
					CT983	Surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem	13/05/2022	13/05/2022
									A070504	Premature Discharge of Battery	13/05/2022	13/05/2022
							F08	Hospitalization or Prolonged Hospitalization	A0705	Battery Problem	13/05/2022	13/05/2022
									A070504	Premature Discharge of Battery	13/05/2022	13/05/2022
							F1905	Device Revision or Replacement	A0705	Battery Problem	13/05/2022	13/05/2022
									A070504	Premature Discharge of Battery	13/05/2022	13/05/2022
6162	2022	All other reportable incidents	37265	Dual-chamber implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	26/04/2022	26/04/2022
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	26/04/2022	26/04/2022
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	26/04/2022	26/04/2022
6163	2022	All other reportable incidents	47265	Dual-chamber implantable pacemaker, rate-responsive	CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem	04/07/2023	04/07/2023
									A070504	Premature Discharge of Battery	04/07/2023	04/07/2023
							F08	Hospitalization or Prolonged Hospitalization	A0705	Battery Problem	04/07/2023	04/07/2023
									A070504	Premature Discharge of Battery	04/07/2023	04/07/2023
							F1905	Device Revision or Replacement	A0705	Battery Problem	04/07/2023	04/07/2023
									A070504	Premature Discharge of Battery	04/07/2023	04/07/2023
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem	04/07/2023	04/07/2023
									A070504	Premature Discharge of Battery	04/07/2023	04/07/2023
							F08	Hospitalization or Prolonged Hospitalization	A0705	Battery Problem	04/07/2023	04/07/2023
									A070504	Premature Discharge of Battery	04/07/2023	04/07/2023
							F1905	Device Revision or Replacement	A0705	Battery Problem	04/07/2023	04/07/2023
									A070504	Premature Discharge of Battery	04/07/2023	04/07/2023
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem	04/07/2023	04/07/2023
									A070504	Premature Discharge of Battery	04/07/2023	04/07/2023
							F08	Hospitalization or Prolonged Hospitalization	A0705	Battery Problem	04/07/2023	04/07/2023
									A070504	Premature Discharge of Battery	04/07/2023	04/07/2023
							F1905	Device Revision or Replacement	A0705	Battery Problem	04/07/2023	04/07/2023
									A070504	Premature Discharge of Battery	04/07/2023	04/07/2023
6164	2022	All other reportable incidents	36241	Pacemaker/defibrillator lead	CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070909	Oversensing		
							F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A070909	Oversensing		
6165	2022	All other reportable incidents	35852	Single-chamber implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A070910	Undersensing		
									A0712	Pacing Problem		
							F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A070910	Undersensing		
									A0712	Pacing Problem		
6166	2021	All other reportable incidents	35852	Single-chamber implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E2104	Electric Shock	A070909	Oversensing		
									A071302	Inappropriate/Inadequate Shock/Stimulation		
							E2403	No Clinical Signs, Symptoms or Conditions	A070909	Oversensing		
									A071302	Inappropriate/Inadequate Shock/Stimulation		
							F26	No Health Consequences or Impact	A070909	Oversensing		
									A071302	Inappropriate/Inadequate Shock/Stimulation		
6167	2021	All other reportable incidents	47265	Dual-chamber implantable pacemaker, rate-responsive	CT141	Electrodes/Leads and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem	01/12/2021	01/12/2021
									A070504	Premature Discharge of Battery	01/12/2021	01/12/2021
							F08	Hospitalization or Prolonged Hospitalization	A0705	Battery Problem	01/12/2021	01/12/2021
									A070504	Premature Discharge of Battery	01/12/2021	01/12/2021
							F1905	Device Revision or Replacement	A0705	Battery Problem	01/12/2021	01/12/2021
									A070504	Premature Discharge of Battery	01/12/2021	01/12/2021
					CT145	Cardiovascular devices	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem	01/12/2021	01/12/2021
									A070504	Premature Discharge of Battery	01/12/2021	01/12/2021
							F08	Hospitalization or Prolonged Hospitalization	A0705	Battery Problem	01/12/2021	01/12/2021
									A070504	Premature Discharge of Battery	01/12/2021	01/12/2021
							F1905	Device Revision or Replacement	A0705	Battery Problem	01/12/2021	01/12/2021
									A070504	Premature Discharge of Battery	01/12/2021	01/12/2021
					CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem	01/12/2021	01/12/2021
									A070504	Premature Discharge of Battery	01/12/2021	01/12/2021
							F08	Hospitalization or Prolonged Hospitalization	A0705	Battery Problem	01/12/2021	01/12/2021
									A070504	Premature Discharge of Battery	01/12/2021	01/12/2021
							F1905	Device Revision or Replacement	A0705	Battery Problem	01/12/2021	01/12/2021
									A070504	Premature Discharge of Battery	01/12/2021	01/12/2021
					CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem	01/12/2021	01/12/2021
									A070504	Premature Discharge of Battery	01/12/2021	01/12/2021
							F08	Hospitalization or Prolonged Hospitalization	A0705	Battery Problem	01/12/2021	01/12/2021
									A070504	Premature Discharge of Battery	01/12/2021	01/12/2021
							F1905	Device Revision or Replacement	A0705	Battery Problem	01/12/2021	01/12/2021

								A070504	Premature Discharge of Battery	01/12/2021	01/12/2021
					CT233	Surgical	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem	01/12/2021	01/12/2021
									A070504	Premature Discharge of Battery	01/12/2021	01/12/2021
							F08	Hospitalization or Prolonged Hospitalization	A0705	Battery Problem	01/12/2021	01/12/2021
									A070504	Premature Discharge of Battery	01/12/2021	01/12/2021
							F1905	Device Revision or Replacement	A0705	Battery Problem	01/12/2021	01/12/2021
									A070504	Premature Discharge of Battery	01/12/2021	01/12/2021
					CT2406	Implantable	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem	01/12/2021	01/12/2021
									A070504	Premature Discharge of Battery	01/12/2021	01/12/2021
							F08	Hospitalization or Prolonged Hospitalization	A0705	Battery Problem	01/12/2021	01/12/2021
									A070504	Premature Discharge of Battery	01/12/2021	01/12/2021
							F1905	Device Revision or Replacement	A0705	Battery Problem	01/12/2021	01/12/2021
									A070504	Premature Discharge of Battery	01/12/2021	01/12/2021
					CT2407	Active implantable	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem	01/12/2021	01/12/2021
									A070504	Premature Discharge of Battery	01/12/2021	01/12/2021
							F08	Hospitalization or Prolonged Hospitalization	A0705	Battery Problem	01/12/2021	01/12/2021
									A070504	Premature Discharge of Battery	01/12/2021	01/12/2021
							F1905	Device Revision or Replacement	A0705	Battery Problem	01/12/2021	01/12/2021
									A070504	Premature Discharge of Battery	01/12/2021	01/12/2021
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem	01/12/2021	01/12/2021
									A070504	Premature Discharge of Battery	01/12/2021	01/12/2021
							F08	Hospitalization or Prolonged Hospitalization	A0705	Battery Problem	01/12/2021	01/12/2021
									A070504	Premature Discharge of Battery	01/12/2021	01/12/2021
							F1905	Device Revision or Replacement	A0705	Battery Problem	01/12/2021	01/12/2021
									A070504	Premature Discharge of Battery	01/12/2021	01/12/2021
					CT2727	Active implant devices/ancillaries	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem	01/12/2021	01/12/2021
									A070504	Premature Discharge of Battery	01/12/2021	01/12/2021
							F08	Hospitalization or Prolonged Hospitalization	A0705	Battery Problem	01/12/2021	01/12/2021
									A070504	Premature Discharge of Battery	01/12/2021	01/12/2021
							F1905	Device Revision or Replacement	A0705	Battery Problem	01/12/2021	01/12/2021
									A070504	Premature Discharge of Battery	01/12/2021	01/12/2021
					CT2750	Implantable cardiac electrical stimulators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem	01/12/2021	01/12/2021
									A070504	Premature Discharge of Battery	01/12/2021	01/12/2021
							F08	Hospitalization or Prolonged Hospitalization	A0705	Battery Problem	01/12/2021	01/12/2021
									A070504	Premature Discharge of Battery	01/12/2021	01/12/2021
							F1905	Device Revision or Replacement	A0705	Battery Problem	01/12/2021	01/12/2021
									A070504	Premature Discharge of Battery	01/12/2021	01/12/2021
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem	01/12/2021	01/12/2021
									A070504	Premature Discharge of Battery	01/12/2021	01/12/2021
							F08	Hospitalization or Prolonged Hospitalization	A0705	Battery Problem	01/12/2021	01/12/2021
									A070504	Premature Discharge of Battery	01/12/2021	01/12/2021
							F1905	Device Revision or Replacement	A0705	Battery Problem	01/12/2021	01/12/2021
									A070504	Premature Discharge of Battery	01/12/2021	01/12/2021
					CT2985	Active	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem	01/12/2021	01/12/2021
									A070504	Premature Discharge of Battery	01/12/2021	01/12/2021
							F08	Hospitalization or Prolonged Hospitalization	A0705	Battery Problem	01/12/2021	01/12/2021
									A070504	Premature Discharge of Battery	01/12/2021	01/12/2021
							F1905	Device Revision or Replacement	A0705	Battery Problem	01/12/2021	01/12/2021
									A070504	Premature Discharge of Battery	01/12/2021	01/12/2021
					CT3001	Implantable metal	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem	01/12/2021	01/12/2021
									A070504	Premature Discharge of Battery	01/12/2021	01/12/2021
							F08	Hospitalization or Prolonged Hospitalization	A0705	Battery Problem	01/12/2021	01/12/2021
									A070504	Premature Discharge of Battery	01/12/2021	01/12/2021
							F1905	Device Revision or Replacement	A0705	Battery Problem	01/12/2021	01/12/2021
									A070504	Premature Discharge of Battery	01/12/2021	01/12/2021
					CT3002	Implantable synthetic polymer	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem	01/12/2021	01/12/2021
									A070504	Premature Discharge of Battery	01/12/2021	01/12/2021
							F08	Hospitalization or Prolonged Hospitalization	A0705	Battery Problem	01/12/2021	01/12/2021
									A070504	Premature Discharge of Battery	01/12/2021	01/12/2021
							F1905	Device Revision or Replacement	A0705	Battery Problem	01/12/2021	01/12/2021
									A070504	Premature Discharge of Battery	01/12/2021	01/12/2021
					CT3006	Implantable inorganic material	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem	01/12/2021	01/12/2021
									A070504	Premature Discharge of Battery	01/12/2021	01/12/2021
							F08	Hospitalization or Prolonged Hospitalization	A0705	Battery Problem	01/12/2021	01/12/2021
									A070504	Premature Discharge of Battery	01/12/2021	01/12/2021
							F1905	Device Revision or Replacement	A0705	Battery Problem	01/12/2021	01/12/2021
									A070504	Premature Discharge of Battery	01/12/2021	01/12/2021
					CT314	Electrically powered	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem	01/12/2021	01/12/2021
									A070504	Premature Discharge of Battery	01/12/2021	01/12/2021
							F08	Hospitalization or Prolonged Hospitalization	A0705	Battery Problem	01/12/2021	01/12/2021
									A070504	Premature Discharge of Battery	01/12/2021	01/12/2021
							F1905	Device Revision or Replacement	A0705	Battery Problem	01/12/2021	01/12/2021
									A070504	Premature Discharge of Battery	01/12/2021	01/12/2021
					CT321	Long-term surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem	01/12/2021	01/12/2021
									A070504	Premature Discharge of Battery	01/12/2021	01/12/2021
							F08	Hospitalization or Prolonged Hospitalization	A0705	Battery Problem	01/12/2021	01/12/2021
									A070504	Premature Discharge of Battery	01/12/2021	01/12/2021
							F1905	Device Revision or Replacement	A0705	Battery Problem	01/12/2021	01/12/2021
									A070504	Premature Discharge of Battery	01/12/2021	01/12/2021
					CT449	Generators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem	01/12/2021	01/12/2021
									A070504	Premature Discharge of Battery	01/12/2021	01/12/2021
							F08	Hospitalization or Prolonged Hospitalization	A0705	Battery Problem	01/12/2021	01/12/2021
									A070504	Premature Discharge of Battery	01/12/2021	01/12/2021
							F1905	Device Revision or Replacement	A0705	Battery Problem	01/12/2021	01/12/2021
									A070504	Premature Discharge of Battery	01/12/2021	01/12/2021
					CT981	Single use	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem	01/12/2021	01/12/2021
									A070504	Premature Discharge of Battery	01/12/2021	01/12/2021
							F08	Hospitalization or Prolonged Hospitalization	A0705	Battery Problem	01/12/2021	01/12/2021

								A070504	Premature Discharge of Battery	01/12/2021	01/12/2021
						F1905	Device Revision or Replacement	A0705	Battery Problem	01/12/2021	01/12/2021
								A070504	Premature Discharge of Battery	01/12/2021	01/12/2021
					CT983	Surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem	01/12/2021	01/12/2021
									A070504	Premature Discharge of Battery	01/12/2021	01/12/2021
							F08	Hospitalization or Prolonged Hospitalization	A0705	Battery Problem	01/12/2021	01/12/2021
									A070504	Premature Discharge of Battery	01/12/2021	01/12/2021
							F1905	Device Revision or Replacement	A0705	Battery Problem	01/12/2021	01/12/2021
									A070504	Premature Discharge of Battery	01/12/2021	01/12/2021
6168	2021	All other reportable incidents	47267	Single-chamber implantable pacemaker, rate-responsive	CT141	Electrodes/Leads and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem	09/08/2021	09/08/2021
									A070504	Premature Discharge of Battery	09/08/2021	09/08/2021
							F08	Hospitalization or Prolonged Hospitalization	A0705	Battery Problem	09/08/2021	09/08/2021
									A070504	Premature Discharge of Battery	09/08/2021	09/08/2021
							F1905	Device Revision or Replacement	A0705	Battery Problem	09/08/2021	09/08/2021
									A070504	Premature Discharge of Battery	09/08/2021	09/08/2021
					CT145	Cardiovascular devices	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem	09/08/2021	09/08/2021
									A070504	Premature Discharge of Battery	09/08/2021	09/08/2021
							F08	Hospitalization or Prolonged Hospitalization	A0705	Battery Problem	09/08/2021	09/08/2021
									A070504	Premature Discharge of Battery	09/08/2021	09/08/2021
							F1905	Device Revision or Replacement	A0705	Battery Problem	09/08/2021	09/08/2021
									A070504	Premature Discharge of Battery	09/08/2021	09/08/2021
					CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem	09/08/2021	09/08/2021
									A070504	Premature Discharge of Battery	09/08/2021	09/08/2021
							F08	Hospitalization or Prolonged Hospitalization	A0705	Battery Problem	09/08/2021	09/08/2021
									A070504	Premature Discharge of Battery	09/08/2021	09/08/2021
							F1905	Device Revision or Replacement	A0705	Battery Problem	09/08/2021	09/08/2021
									A070504	Premature Discharge of Battery	09/08/2021	09/08/2021
					CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem	09/08/2021	09/08/2021
									A070504	Premature Discharge of Battery	09/08/2021	09/08/2021
							F08	Hospitalization or Prolonged Hospitalization	A0705	Battery Problem	09/08/2021	09/08/2021
									A070504	Premature Discharge of Battery	09/08/2021	09/08/2021
							F1905	Device Revision or Replacement	A0705	Battery Problem	09/08/2021	09/08/2021
									A070504	Premature Discharge of Battery	09/08/2021	09/08/2021
					CT233	Surgical	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem	09/08/2021	09/08/2021
									A070504	Premature Discharge of Battery	09/08/2021	09/08/2021
							F08	Hospitalization or Prolonged Hospitalization	A0705	Battery Problem	09/08/2021	09/08/2021
									A070504	Premature Discharge of Battery	09/08/2021	09/08/2021
							F1905	Device Revision or Replacement	A0705	Battery Problem	09/08/2021	09/08/2021
									A070504	Premature Discharge of Battery	09/08/2021	09/08/2021
					CT2406	Implantable	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem	09/08/2021	09/08/2021
									A070504	Premature Discharge of Battery	09/08/2021	09/08/2021
							F08	Hospitalization or Prolonged Hospitalization	A0705	Battery Problem	09/08/2021	09/08/2021
									A070504	Premature Discharge of Battery	09/08/2021	09/08/2021
							F1905	Device Revision or Replacement	A0705	Battery Problem	09/08/2021	09/08/2021
									A070504	Premature Discharge of Battery	09/08/2021	09/08/2021
					CT2407	Active implantable	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem	09/08/2021	09/08/2021
									A070504	Premature Discharge of Battery	09/08/2021	09/08/2021
							F08	Hospitalization or Prolonged Hospitalization	A0705	Battery Problem	09/08/2021	09/08/2021
									A070504	Premature Discharge of Battery	09/08/2021	09/08/2021
							F1905	Device Revision or Replacement	A0705	Battery Problem	09/08/2021	09/08/2021
									A070504	Premature Discharge of Battery	09/08/2021	09/08/2021
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem	09/08/2021	09/08/2021
									A070504	Premature Discharge of Battery	09/08/2021	09/08/2021
							F08	Hospitalization or Prolonged Hospitalization	A0705	Battery Problem	09/08/2021	09/08/2021
									A070504	Premature Discharge of Battery	09/08/2021	09/08/2021
							F1905	Device Revision or Replacement	A0705	Battery Problem	09/08/2021	09/08/2021
									A070504	Premature Discharge of Battery	09/08/2021	09/08/2021
					CT2727	Active implant devices/ancillaries	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem	09/08/2021	09/08/2021
									A070504	Premature Discharge of Battery	09/08/2021	09/08/2021
							F08	Hospitalization or Prolonged Hospitalization	A0705	Battery Problem	09/08/2021	09/08/2021
									A070504	Premature Discharge of Battery	09/08/2021	09/08/2021
							F1905	Device Revision or Replacement	A0705	Battery Problem	09/08/2021	09/08/2021
									A070504	Premature Discharge of Battery	09/08/2021	09/08/2021
					CT2750	Implantable cardiac electrical stimulators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem	09/08/2021	09/08/2021
									A070504	Premature Discharge of Battery	09/08/2021	09/08/2021
							F08	Hospitalization or Prolonged Hospitalization	A0705	Battery Problem	09/08/2021	09/08/2021
									A070504	Premature Discharge of Battery	09/08/2021	09/08/2021
							F1905	Device Revision or Replacement	A0705	Battery Problem	09/08/2021	09/08/2021
									A070504	Premature Discharge of Battery	09/08/2021	09/08/2021
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem	09/08/2021	09/08/2021
									A070504	Premature Discharge of Battery	09/08/2021	09/08/2021
							F08	Hospitalization or Prolonged Hospitalization	A0705	Battery Problem	09/08/2021	09/08/2021
									A070504	Premature Discharge of Battery	09/08/2021	09/08/2021
							F1905	Device Revision or Replacement	A0705	Battery Problem	09/08/2021	09/08/2021
									A070504	Premature Discharge of Battery	09/08/2021	09/08/2021
					CT2985	Active	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem	09/08/2021	09/08/2021
									A070504	Premature Discharge of Battery	09/08/2021	09/08/2021
							F08	Hospitalization or Prolonged Hospitalization	A0705	Battery Problem	09/08/2021	09/08/2021
									A070504	Premature Discharge of Battery	09/08/2021	09/08/2021
							F1905	Device Revision or Replacement	A0705	Battery Problem	09/08/2021	09/08/2021
									A070504	Premature Discharge of Battery	09/08/2021	09/08/2021
					CT3001	Implantable metal	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem	09/08/2021	09/08/2021
									A070504	Premature Discharge of Battery	09/08/2021	09/08/2021
							F08	Hospitalization or Prolonged Hospitalization	A0705	Battery Problem	09/08/2021	09/08/2021
									A070504	Premature Discharge of Battery	09/08/2021	09/08/2021
							F1905	Device Revision or Replacement	A0705	Battery Problem	09/08/2021	09/08/2021
									A070504	Premature Discharge of Battery	09/08/2021	09/08/2021
					CT3002	Implantable synthetic polymer	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem	09/08/2021	09/08/2021

									A070504	Premature Discharge of Battery	09/08/2021	09/08/2021
							F08	Hospitalization or Prolonged Hospitalization	A0705	Battery Problem	09/08/2021	09/08/2021
									A070504	Premature Discharge of Battery	09/08/2021	09/08/2021
							F1905	Device Revision or Replacement	A0705	Battery Problem	09/08/2021	09/08/2021
									A070504	Premature Discharge of Battery	09/08/2021	09/08/2021
					CT3006	Implantable inorganic material	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem	09/08/2021	09/08/2021
									A070504	Premature Discharge of Battery	09/08/2021	09/08/2021
							F08	Hospitalization or Prolonged Hospitalization	A0705	Battery Problem	09/08/2021	09/08/2021
									A070504	Premature Discharge of Battery	09/08/2021	09/08/2021
							F1905	Device Revision or Replacement	A0705	Battery Problem	09/08/2021	09/08/2021
									A070504	Premature Discharge of Battery	09/08/2021	09/08/2021
					CT314	Electrically powered	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem	09/08/2021	09/08/2021
									A070504	Premature Discharge of Battery	09/08/2021	09/08/2021
							F08	Hospitalization or Prolonged Hospitalization	A0705	Battery Problem	09/08/2021	09/08/2021
									A070504	Premature Discharge of Battery	09/08/2021	09/08/2021
							F1905	Device Revision or Replacement	A0705	Battery Problem	09/08/2021	09/08/2021
									A070504	Premature Discharge of Battery	09/08/2021	09/08/2021
					CT321	Long-term surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem	09/08/2021	09/08/2021
									A070504	Premature Discharge of Battery	09/08/2021	09/08/2021
							F08	Hospitalization or Prolonged Hospitalization	A0705	Battery Problem	09/08/2021	09/08/2021
									A070504	Premature Discharge of Battery	09/08/2021	09/08/2021
							F1905	Device Revision or Replacement	A0705	Battery Problem	09/08/2021	09/08/2021
									A070504	Premature Discharge of Battery	09/08/2021	09/08/2021
					CT449	Generators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem	09/08/2021	09/08/2021
									A070504	Premature Discharge of Battery	09/08/2021	09/08/2021
							F08	Hospitalization or Prolonged Hospitalization	A0705	Battery Problem	09/08/2021	09/08/2021
									A070504	Premature Discharge of Battery	09/08/2021	09/08/2021
							F1905	Device Revision or Replacement	A0705	Battery Problem	09/08/2021	09/08/2021
									A070504	Premature Discharge of Battery	09/08/2021	09/08/2021
					CT981	Single use	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem	09/08/2021	09/08/2021
									A070504	Premature Discharge of Battery	09/08/2021	09/08/2021
							F08	Hospitalization or Prolonged Hospitalization	A0705	Battery Problem	09/08/2021	09/08/2021
									A070504	Premature Discharge of Battery	09/08/2021	09/08/2021
							F1905	Device Revision or Replacement	A0705	Battery Problem	09/08/2021	09/08/2021
									A070504	Premature Discharge of Battery	09/08/2021	09/08/2021
					CT983	Surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem	09/08/2021	09/08/2021
									A070504	Premature Discharge of Battery	09/08/2021	09/08/2021
							F08	Hospitalization or Prolonged Hospitalization	A0705	Battery Problem	09/08/2021	09/08/2021
									A070504	Premature Discharge of Battery	09/08/2021	09/08/2021
							F1905	Device Revision or Replacement	A0705	Battery Problem	09/08/2021	09/08/2021
									A070504	Premature Discharge of Battery	09/08/2021	09/08/2021
6169	2021	All other reportable incidents	36241	Pacemaker/defibrillator lead	CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A040101	Fracture	26/09/2021	26/09/2021
									A070909	Oversensing	26/09/2021	26/09/2021
							F08	Hospitalization or Prolonged Hospitalization	A040101	Fracture	26/09/2021	26/09/2021
									A070909	Oversensing	26/09/2021	26/09/2021
							F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A040101	Fracture	26/09/2021	26/09/2021
									A070909	Oversensing	26/09/2021	26/09/2021
							F1905	Device Revision or Replacement	A040101	Fracture	26/09/2021	26/09/2021
									A070909	Oversensing	26/09/2021	26/09/2021
6170	2021	Unanticipated serious deterioration in state of health	47270	Cardiac resynchronization therapy implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E011903	Syncope/Fainting	A070101	Failure to Capture	26/07/2021	26/07/2021
									A0712	Pacing Problem	26/07/2021	26/07/2021
									A13	Communication or Transmission Problem	26/07/2021	26/07/2021
							F12	Serious Injury/ Illness/ Impairment	A070101	Failure to Capture	26/07/2021	26/07/2021
									A0712	Pacing Problem	26/07/2021	26/07/2021
									A13	Communication or Transmission Problem	26/07/2021	26/07/2021
							F1905	Device Revision or Replacement	A070101	Failure to Capture	26/07/2021	26/07/2021
									A0712	Pacing Problem	26/07/2021	26/07/2021
									A13	Communication or Transmission Problem	26/07/2021	26/07/2021
6171	2021	All other reportable incidents	35852	Single-chamber implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem	30/07/2021	30/07/2021
									A070504	Premature Discharge of Battery	30/07/2021	30/07/2021
							F08	Hospitalization or Prolonged Hospitalization	A0705	Battery Problem	30/07/2021	30/07/2021
									A070504	Premature Discharge of Battery	30/07/2021	30/07/2021
							F1905	Device Revision or Replacement	A0705	Battery Problem	30/07/2021	30/07/2021
									A070504	Premature Discharge of Battery	30/07/2021	30/07/2021
6172	2021	All other reportable incidents	35223	Endocardial/interventricular septal pacing lead	CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A040101	Fracture		
							F26	No Health Consequences or Impact	A040101	Fracture		
6173	2022	All other reportable incidents	36241	Pacemaker/defibrillator lead	CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A051201	Device Dislodged or Dislocated		
							F08	Hospitalization or Prolonged Hospitalization	A051201	Device Dislodged or Dislocated		
							F1904	Device Repositioning	A051201	Device Dislodged or Dislocated		
6174	2022	All other reportable incidents	47265	Dual-chamber implantable pacemaker, rate-responsive	CT141	Electrodes/Leads and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem	31/05/2022	31/05/2022
									A070504	Premature Discharge of Battery	31/05/2022	31/05/2022
							F08	Hospitalization or Prolonged Hospitalization	A0705	Battery Problem	31/05/2022	31/05/2022
									A070504	Premature Discharge of Battery	31/05/2022	31/05/2022
							F1905	Device Revision or Replacement	A0705	Battery Problem	31/05/2022	31/05/2022
									A070504	Premature Discharge of Battery	31/05/2022	31/05/2022
					CT145	Cardiovascular devices	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem	31/05/2022	31/05/2022
									A070504	Premature Discharge of Battery	31/05/2022	31/05/2022
							F08	Hospitalization or Prolonged Hospitalization	A0705	Battery Problem	31/05/2022	31/05/2022
									A070504	Premature Discharge of Battery	31/05/2022	31/05/2022
							F1905	Device Revision or Replacement	A0705	Battery Problem	31/05/2022	31/05/2022
									A070504	Premature Discharge of Battery	31/05/2022	31/05/2022
					CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem	31/05/2022	31/05/2022
									A070504	Premature Discharge of Battery	31/05/2022	31/05/2022
							F08	Hospitalization or Prolonged Hospitalization	A0705	Battery Problem	31/05/2022	31/05/2022
									A070504	Premature Discharge of Battery	31/05/2022	31/05/2022
							F1905	Device Revision or Replacement	A0705	Battery Problem	31/05/2022	31/05/2022
									A070504	Premature Discharge of Battery	31/05/2022	31/05/2022
					CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem	31/05/2022	31/05/2022

									A070504	Premature Discharge of Battery	31/05/2022	31/05/2022
							F08	Hospitalization or Prolonged Hospitalization	A0705	Battery Problem	31/05/2022	31/05/2022
									A070504	Premature Discharge of Battery	31/05/2022	31/05/2022
							F1905	Device Revision or Replacement	A0705	Battery Problem	31/05/2022	31/05/2022
									A070504	Premature Discharge of Battery	31/05/2022	31/05/2022
					CT233	Surgical	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem	31/05/2022	31/05/2022
									A070504	Premature Discharge of Battery	31/05/2022	31/05/2022
							F08	Hospitalization or Prolonged Hospitalization	A0705	Battery Problem	31/05/2022	31/05/2022
									A070504	Premature Discharge of Battery	31/05/2022	31/05/2022
							F1905	Device Revision or Replacement	A0705	Battery Problem	31/05/2022	31/05/2022
									A070504	Premature Discharge of Battery	31/05/2022	31/05/2022
					CT2406	Implantable	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem	31/05/2022	31/05/2022
									A070504	Premature Discharge of Battery	31/05/2022	31/05/2022
							F08	Hospitalization or Prolonged Hospitalization	A0705	Battery Problem	31/05/2022	31/05/2022
									A070504	Premature Discharge of Battery	31/05/2022	31/05/2022
							F1905	Device Revision or Replacement	A0705	Battery Problem	31/05/2022	31/05/2022
									A070504	Premature Discharge of Battery	31/05/2022	31/05/2022
					CT2407	Active implantable	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem	31/05/2022	31/05/2022
									A070504	Premature Discharge of Battery	31/05/2022	31/05/2022
							F08	Hospitalization or Prolonged Hospitalization	A0705	Battery Problem	31/05/2022	31/05/2022
									A070504	Premature Discharge of Battery	31/05/2022	31/05/2022
							F1905	Device Revision or Replacement	A0705	Battery Problem	31/05/2022	31/05/2022
									A070504	Premature Discharge of Battery	31/05/2022	31/05/2022
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem	31/05/2022	31/05/2022
									A070504	Premature Discharge of Battery	31/05/2022	31/05/2022
							F08	Hospitalization or Prolonged Hospitalization	A0705	Battery Problem	31/05/2022	31/05/2022
									A070504	Premature Discharge of Battery	31/05/2022	31/05/2022
							F1905	Device Revision or Replacement	A0705	Battery Problem	31/05/2022	31/05/2022
									A070504	Premature Discharge of Battery	31/05/2022	31/05/2022
					CT2727	Active implant devices/ancillaries	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem	31/05/2022	31/05/2022
									A070504	Premature Discharge of Battery	31/05/2022	31/05/2022
							F08	Hospitalization or Prolonged Hospitalization	A0705	Battery Problem	31/05/2022	31/05/2022
									A070504	Premature Discharge of Battery	31/05/2022	31/05/2022
							F1905	Device Revision or Replacement	A0705	Battery Problem	31/05/2022	31/05/2022
									A070504	Premature Discharge of Battery	31/05/2022	31/05/2022
					CT2750	Implantable cardiac electrical stimulators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem	31/05/2022	31/05/2022
									A070504	Premature Discharge of Battery	31/05/2022	31/05/2022
							F08	Hospitalization or Prolonged Hospitalization	A0705	Battery Problem	31/05/2022	31/05/2022
									A070504	Premature Discharge of Battery	31/05/2022	31/05/2022
							F1905	Device Revision or Replacement	A0705	Battery Problem	31/05/2022	31/05/2022
									A070504	Premature Discharge of Battery	31/05/2022	31/05/2022
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem	31/05/2022	31/05/2022
									A070504	Premature Discharge of Battery	31/05/2022	31/05/2022
							F08	Hospitalization or Prolonged Hospitalization	A0705	Battery Problem	31/05/2022	31/05/2022
									A070504	Premature Discharge of Battery	31/05/2022	31/05/2022
							F1905	Device Revision or Replacement	A0705	Battery Problem	31/05/2022	31/05/2022
									A070504	Premature Discharge of Battery	31/05/2022	31/05/2022
					CT2985	Active	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem	31/05/2022	31/05/2022
									A070504	Premature Discharge of Battery	31/05/2022	31/05/2022
							F08	Hospitalization or Prolonged Hospitalization	A0705	Battery Problem	31/05/2022	31/05/2022
									A070504	Premature Discharge of Battery	31/05/2022	31/05/2022
							F1905	Device Revision or Replacement	A0705	Battery Problem	31/05/2022	31/05/2022
									A070504	Premature Discharge of Battery	31/05/2022	31/05/2022
					CT3001	Implantable metal	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem	31/05/2022	31/05/2022
									A070504	Premature Discharge of Battery	31/05/2022	31/05/2022
							F08	Hospitalization or Prolonged Hospitalization	A0705	Battery Problem	31/05/2022	31/05/2022
									A070504	Premature Discharge of Battery	31/05/2022	31/05/2022
							F1905	Device Revision or Replacement	A0705	Battery Problem	31/05/2022	31/05/2022
									A070504	Premature Discharge of Battery	31/05/2022	31/05/2022
					CT3002	Implantable synthetic polymer	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem	31/05/2022	31/05/2022
									A070504	Premature Discharge of Battery	31/05/2022	31/05/2022
							F08	Hospitalization or Prolonged Hospitalization	A0705	Battery Problem	31/05/2022	31/05/2022
									A070504	Premature Discharge of Battery	31/05/2022	31/05/2022
							F1905	Device Revision or Replacement	A0705	Battery Problem	31/05/2022	31/05/2022
									A070504	Premature Discharge of Battery	31/05/2022	31/05/2022
					CT3006	Implantable inorganic material	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem	31/05/2022	31/05/2022
									A070504	Premature Discharge of Battery	31/05/2022	31/05/2022
							F08	Hospitalization or Prolonged Hospitalization	A0705	Battery Problem	31/05/2022	31/05/2022
									A070504	Premature Discharge of Battery	31/05/2022	31/05/2022
							F1905	Device Revision or Replacement	A0705	Battery Problem	31/05/2022	31/05/2022
									A070504	Premature Discharge of Battery	31/05/2022	31/05/2022
					CT314	Electrically powered	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem	31/05/2022	31/05/2022
									A070504	Premature Discharge of Battery	31/05/2022	31/05/2022
							F08	Hospitalization or Prolonged Hospitalization	A0705	Battery Problem	31/05/2022	31/05/2022
									A070504	Premature Discharge of Battery	31/05/2022	31/05/2022
							F1905	Device Revision or Replacement	A0705	Battery Problem	31/05/2022	31/05/2022
									A070504	Premature Discharge of Battery	31/05/2022	31/05/2022
					CT321	Long-term surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem	31/05/2022	31/05/2022
									A070504	Premature Discharge of Battery	31/05/2022	31/05/2022
							F08	Hospitalization or Prolonged Hospitalization	A0705	Battery Problem	31/05/2022	31/05/2022
									A070504	Premature Discharge of Battery	31/05/2022	31/05/2022
							F1905	Device Revision or Replacement	A0705	Battery Problem	31/05/2022	31/05/2022
									A070504	Premature Discharge of Battery	31/05/2022	31/05/2022
					CT449	Generators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem	31/05/2022	31/05/2022
									A070504	Premature Discharge of Battery	31/05/2022	31/05/2022
							F08	Hospitalization or Prolonged Hospitalization	A0705	Battery Problem	31/05/2022	31/05/2022
									A070504	Premature Discharge of Battery	31/05/2022	31/05/2022
							F1905	Device Revision or Replacement	A0705	Battery Problem	31/05/2022	31/05/2022

									A070504	Premature Discharge of Battery	31/05/2022	31/05/2022
					CT981	Single use	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem	31/05/2022	31/05/2022
									A070504	Premature Discharge of Battery	31/05/2022	31/05/2022
							F08	Hospitalization or Prolonged Hospitalization	A0705	Battery Problem	31/05/2022	31/05/2022
									A070504	Premature Discharge of Battery	31/05/2022	31/05/2022
							F1905	Device Revision or Replacement	A0705	Battery Problem	31/05/2022	31/05/2022
									A070504	Premature Discharge of Battery	31/05/2022	31/05/2022
					CT983	Surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem	31/05/2022	31/05/2022
									A070504	Premature Discharge of Battery	31/05/2022	31/05/2022
							F08	Hospitalization or Prolonged Hospitalization	A0705	Battery Problem	31/05/2022	31/05/2022
									A070504	Premature Discharge of Battery	31/05/2022	31/05/2022
							F1905	Device Revision or Replacement	A0705	Battery Problem	31/05/2022	31/05/2022
									A070504	Premature Discharge of Battery	31/05/2022	31/05/2022
6175	2022	All other reportable incidents	47267	Single-chamber implantable pacemaker, rate-responsive	CT141	Electrodes/Leads and associated devices	E0112	Dizziness	A070101	Failure to Capture	01/02/2022	28/02/2022
							E2403	No Clinical Signs, Symptoms or Conditions	A070101	Failure to Capture	01/02/2022	28/02/2022
							F08	Hospitalization or Prolonged Hospitalization	A070101	Failure to Capture	01/02/2022	28/02/2022
							F19	Surgical Intervention	A070101	Failure to Capture	01/02/2022	28/02/2022
					CT145	Cardiovascular devices	E0112	Dizziness	A070101	Failure to Capture	01/02/2022	28/02/2022
							E2403	No Clinical Signs, Symptoms or Conditions	A070101	Failure to Capture	01/02/2022	28/02/2022
							F08	Hospitalization or Prolonged Hospitalization	A070101	Failure to Capture	01/02/2022	28/02/2022
							F19	Surgical Intervention	A070101	Failure to Capture	01/02/2022	28/02/2022
					CT1684	Cardiac pacemakers and associated devices	E0112	Dizziness	A070101	Failure to Capture	01/02/2022	28/02/2022
							E2403	No Clinical Signs, Symptoms or Conditions	A070101	Failure to Capture	01/02/2022	28/02/2022
							F08	Hospitalization or Prolonged Hospitalization	A070101	Failure to Capture	01/02/2022	28/02/2022
							F19	Surgical Intervention	A070101	Failure to Capture	01/02/2022	28/02/2022
					CT1685	Cardiac pacemakers	E0112	Dizziness	A070101	Failure to Capture	01/02/2022	28/02/2022
							E2403	No Clinical Signs, Symptoms or Conditions	A070101	Failure to Capture	01/02/2022	28/02/2022
							F08	Hospitalization or Prolonged Hospitalization	A070101	Failure to Capture	01/02/2022	28/02/2022
							F19	Surgical Intervention	A070101	Failure to Capture	01/02/2022	28/02/2022
					CT233	Surgical	E0112	Dizziness	A070101	Failure to Capture	01/02/2022	28/02/2022
							E2403	No Clinical Signs, Symptoms or Conditions	A070101	Failure to Capture	01/02/2022	28/02/2022
							F08	Hospitalization or Prolonged Hospitalization	A070101	Failure to Capture	01/02/2022	28/02/2022
							F19	Surgical Intervention	A070101	Failure to Capture	01/02/2022	28/02/2022
					CT2406	Implantable	E0112	Dizziness	A070101	Failure to Capture	01/02/2022	28/02/2022
							E2403	No Clinical Signs, Symptoms or Conditions	A070101	Failure to Capture	01/02/2022	28/02/2022
							F08	Hospitalization or Prolonged Hospitalization	A070101	Failure to Capture	01/02/2022	28/02/2022
							F19	Surgical Intervention	A070101	Failure to Capture	01/02/2022	28/02/2022
					CT2407	Active implantable	E0112	Dizziness	A070101	Failure to Capture	01/02/2022	28/02/2022
							E2403	No Clinical Signs, Symptoms or Conditions	A070101	Failure to Capture	01/02/2022	28/02/2022
							F08	Hospitalization or Prolonged Hospitalization	A070101	Failure to Capture	01/02/2022	28/02/2022
							F19	Surgical Intervention	A070101	Failure to Capture	01/02/2022	28/02/2022
					CT257	Implantable cardiac pacemakers	E0112	Dizziness	A070101	Failure to Capture	01/02/2022	28/02/2022
							E2403	No Clinical Signs, Symptoms or Conditions	A070101	Failure to Capture	01/02/2022	28/02/2022
							F08	Hospitalization or Prolonged Hospitalization	A070101	Failure to Capture	01/02/2022	28/02/2022
							F19	Surgical Intervention	A070101	Failure to Capture	01/02/2022	28/02/2022
					CT2727	Active implant devices/ancillaries	E0112	Dizziness	A070101	Failure to Capture	01/02/2022	28/02/2022
							E2403	No Clinical Signs, Symptoms or Conditions	A070101	Failure to Capture	01/02/2022	28/02/2022
							F08	Hospitalization or Prolonged Hospitalization	A070101	Failure to Capture	01/02/2022	28/02/2022
							F19	Surgical Intervention	A070101	Failure to Capture	01/02/2022	28/02/2022
					CT2750	Implantable cardiac electrical stimulators and associated devices	E0112	Dizziness	A070101	Failure to Capture	01/02/2022	28/02/2022
							E2403	No Clinical Signs, Symptoms or Conditions	A070101	Failure to Capture	01/02/2022	28/02/2022
							F08	Hospitalization or Prolonged Hospitalization	A070101	Failure to Capture	01/02/2022	28/02/2022
							F19	Surgical Intervention	A070101	Failure to Capture	01/02/2022	28/02/2022
					CT2866	Implantable cardiac electrical stimulators	E0112	Dizziness	A070101	Failure to Capture	01/02/2022	28/02/2022
							E2403	No Clinical Signs, Symptoms or Conditions	A070101	Failure to Capture	01/02/2022	28/02/2022
							F08	Hospitalization or Prolonged Hospitalization	A070101	Failure to Capture	01/02/2022	28/02/2022
							F19	Surgical Intervention	A070101	Failure to Capture	01/02/2022	28/02/2022
					CT2985	Active	E0112	Dizziness	A070101	Failure to Capture	01/02/2022	28/02/2022
							E2403	No Clinical Signs, Symptoms or Conditions	A070101	Failure to Capture	01/02/2022	28/02/2022
							F08	Hospitalization or Prolonged Hospitalization	A070101	Failure to Capture	01/02/2022	28/02/2022
							F19	Surgical Intervention	A070101	Failure to Capture	01/02/2022	28/02/2022
					CT3001	Implantable metal	E0112	Dizziness	A070101	Failure to Capture	01/02/2022	28/02/2022
							E2403	No Clinical Signs, Symptoms or Conditions	A070101	Failure to Capture	01/02/2022	28/02/2022
							F08	Hospitalization or Prolonged Hospitalization	A070101	Failure to Capture	01/02/2022	28/02/2022
							F19	Surgical Intervention	A070101	Failure to Capture	01/02/2022	28/02/2022
					CT3002	Implantable synthetic polymer	E0112	Dizziness	A070101	Failure to Capture	01/02/2022	28/02/2022
							E2403	No Clinical Signs, Symptoms or Conditions	A070101	Failure to Capture	01/02/2022	28/02/2022
							F08	Hospitalization or Prolonged Hospitalization	A070101	Failure to Capture	01/02/2022	28/02/2022
							F19	Surgical Intervention	A070101	Failure to Capture	01/02/2022	28/02/2022
					CT3006	Implantable inorganic material	E0112	Dizziness	A070101	Failure to Capture	01/02/2022	28/02/2022
							E2403	No Clinical Signs, Symptoms or Conditions	A070101	Failure to Capture	01/02/2022	28/02/2022
							F08	Hospitalization or Prolonged Hospitalization	A070101	Failure to Capture	01/02/2022	28/02/2022
							F19	Surgical Intervention	A070101	Failure to Capture	01/02/2022	28/02/2022
					CT314	Electrically powered	E0112	Dizziness	A070101	Failure to Capture	01/02/2022	28/02/2022
							E2403	No Clinical Signs, Symptoms or Conditions	A070101	Failure to Capture	01/02/2022	28/02/2022
							F08	Hospitalization or Prolonged Hospitalization	A070101	Failure to Capture	01/02/2022	28/02/2022
							F19	Surgical Intervention	A070101	Failure to Capture	01/02/2022	28/02/2022
					CT321	Long-term surgical invasive	E0112	Dizziness	A070101	Failure to Capture	01/02/2022	28/02/2022
							E2403	No Clinical Signs, Symptoms or Conditions	A070101	Failure to Capture	01/02/2022	28/02/2022
							F08	Hospitalization or Prolonged Hospitalization	A070101	Failure to Capture	01/02/2022	28/02/2022
							F19	Surgical Intervention	A070101	Failure to Capture	01/02/2022	28/02/2022
					CT449	Generators and associated devices	E0112	Dizziness	A070101	Failure to Capture	01/02/2022	28/02/2022
							E2403	No Clinical Signs, Symptoms or Conditions	A070101	Failure to Capture	01/02/2022	28/02/2022
							F08	Hospitalization or Prolonged Hospitalization	A070101	Failure to Capture	01/02/2022	28/02/2022
							F19	Surgical Intervention	A070101	Failure to Capture	01/02/2022	28/02/2022
					CT981	Single use	E0112	Dizziness	A070101	Failure to Capture	01/02/2022	28/02/2022
							E2403	No Clinical Signs, Symptoms or Conditions	A070101	Failure to Capture	01/02/2022	28/02/2022
							F08	Hospitalization or Prolonged Hospitalization	A070101	Failure to Capture	01/02/2022	28/02/2022

						F19	Surgical Intervention	A070101	Failure to Capture	01/02/2022	28/02/2022
					CT983	Surgical invasive	E0112	Dizziness	A070101	Failure to Capture	01/02/2022	28/02/2022
							E2403	No Clinical Signs, Symptoms or Conditions	A070101	Failure to Capture	01/02/2022	28/02/2022
							F08	Hospitalization or Prolonged Hospitalization	A070101	Failure to Capture	01/02/2022	28/02/2022
							F19	Surgical Intervention	A070101	Failure to Capture	01/02/2022	28/02/2022
6176	2022	All other reportable incidents	47265	Dual-chamber implantable pacemaker, rate-responsive	CT141	Electrodes/Leads and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A072201	High Impedance	10/03/2022	10/03/2022
									A12	Connection Problem	10/03/2022	10/03/2022
							F08	Hospitalization or Prolonged Hospitalization	A072201	High Impedance	10/03/2022	10/03/2022
									A12	Connection Problem	10/03/2022	10/03/2022
							F1905	Device Revision or Replacement	A072201	High Impedance	10/03/2022	10/03/2022
									A12	Connection Problem	10/03/2022	10/03/2022
							F2203	Imaging Required	A072201	High Impedance	10/03/2022	10/03/2022
									A12	Connection Problem	10/03/2022	10/03/2022
					CT145	Cardiovascular devices	E2403	No Clinical Signs, Symptoms or Conditions	A072201	High Impedance	10/03/2022	10/03/2022
									A12	Connection Problem	10/03/2022	10/03/2022
							F08	Hospitalization or Prolonged Hospitalization	A072201	High Impedance	10/03/2022	10/03/2022
									A12	Connection Problem	10/03/2022	10/03/2022
							F1905	Device Revision or Replacement	A072201	High Impedance	10/03/2022	10/03/2022
									A12	Connection Problem	10/03/2022	10/03/2022
							F2203	Imaging Required	A072201	High Impedance	10/03/2022	10/03/2022
									A12	Connection Problem	10/03/2022	10/03/2022
					CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A072201	High Impedance	10/03/2022	10/03/2022
									A12	Connection Problem	10/03/2022	10/03/2022
							F08	Hospitalization or Prolonged Hospitalization	A072201	High Impedance	10/03/2022	10/03/2022
									A12	Connection Problem	10/03/2022	10/03/2022
							F1905	Device Revision or Replacement	A072201	High Impedance	10/03/2022	10/03/2022
									A12	Connection Problem	10/03/2022	10/03/2022
							F2203	Imaging Required	A072201	High Impedance	10/03/2022	10/03/2022
									A12	Connection Problem	10/03/2022	10/03/2022
					CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A072201	High Impedance	10/03/2022	10/03/2022
									A12	Connection Problem	10/03/2022	10/03/2022
							F08	Hospitalization or Prolonged Hospitalization	A072201	High Impedance	10/03/2022	10/03/2022
									A12	Connection Problem	10/03/2022	10/03/2022
							F1905	Device Revision or Replacement	A072201	High Impedance	10/03/2022	10/03/2022
									A12	Connection Problem	10/03/2022	10/03/2022
							F2203	Imaging Required	A072201	High Impedance	10/03/2022	10/03/2022
									A12	Connection Problem	10/03/2022	10/03/2022
					CT233	Surgical	E2403	No Clinical Signs, Symptoms or Conditions	A072201	High Impedance	10/03/2022	10/03/2022
									A12	Connection Problem	10/03/2022	10/03/2022
							F08	Hospitalization or Prolonged Hospitalization	A072201	High Impedance	10/03/2022	10/03/2022
									A12	Connection Problem	10/03/2022	10/03/2022
							F1905	Device Revision or Replacement	A072201	High Impedance	10/03/2022	10/03/2022
									A12	Connection Problem	10/03/2022	10/03/2022
							F2203	Imaging Required	A072201	High Impedance	10/03/2022	10/03/2022
									A12	Connection Problem	10/03/2022	10/03/2022
					CT2406	Implantable	E2403	No Clinical Signs, Symptoms or Conditions	A072201	High Impedance	10/03/2022	10/03/2022
									A12	Connection Problem	10/03/2022	10/03/2022
							F08	Hospitalization or Prolonged Hospitalization	A072201	High Impedance	10/03/2022	10/03/2022
									A12	Connection Problem	10/03/2022	10/03/2022
							F1905	Device Revision or Replacement	A072201	High Impedance	10/03/2022	10/03/2022
									A12	Connection Problem	10/03/2022	10/03/2022
							F2203	Imaging Required	A072201	High Impedance	10/03/2022	10/03/2022
									A12	Connection Problem	10/03/2022	10/03/2022
					CT2407	Active implantable	E2403	No Clinical Signs, Symptoms or Conditions	A072201	High Impedance	10/03/2022	10/03/2022
									A12	Connection Problem	10/03/2022	10/03/2022
							F08	Hospitalization or Prolonged Hospitalization	A072201	High Impedance	10/03/2022	10/03/2022
									A12	Connection Problem	10/03/2022	10/03/2022
							F1905	Device Revision or Replacement	A072201	High Impedance	10/03/2022	10/03/2022
									A12	Connection Problem	10/03/2022	10/03/2022
							F2203	Imaging Required	A072201	High Impedance	10/03/2022	10/03/2022
									A12	Connection Problem	10/03/2022	10/03/2022
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A072201	High Impedance	10/03/2022	10/03/2022
									A12	Connection Problem	10/03/2022	10/03/2022
							F08	Hospitalization or Prolonged Hospitalization	A072201	High Impedance	10/03/2022	10/03/2022
									A12	Connection Problem	10/03/2022	10/03/2022
							F1905	Device Revision or Replacement	A072201	High Impedance	10/03/2022	10/03/2022
									A12	Connection Problem	10/03/2022	10/03/2022
							F2203	Imaging Required	A072201	High Impedance	10/03/2022	10/03/2022
									A12	Connection Problem	10/03/2022	10/03/2022
					CT2727	Active implant devices/ancillaries	E2403	No Clinical Signs, Symptoms or Conditions	A072201	High Impedance	10/03/2022	10/03/2022
									A12	Connection Problem	10/03/2022	10/03/2022
							F08	Hospitalization or Prolonged Hospitalization	A072201	High Impedance	10/03/2022	10/03/2022
									A12	Connection Problem	10/03/2022	10/03/2022
							F1905	Device Revision or Replacement	A072201	High Impedance	10/03/2022	10/03/2022
									A12	Connection Problem	10/03/2022	10/03/2022
							F2203	Imaging Required	A072201	High Impedance	10/03/2022	10/03/2022
									A12	Connection Problem	10/03/2022	10/03/2022
					CT2750	Implantable cardiac electrical stimulators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A072201	High Impedance	10/03/2022	10/03/2022
									A12	Connection Problem	10/03/2022	10/03/2022
							F08	Hospitalization or Prolonged Hospitalization	A072201	High Impedance	10/03/2022	10/03/2022
									A12	Connection Problem	10/03/2022	10/03/2022
							F1905	Device Revision or Replacement	A072201	High Impedance	10/03/2022	10/03/2022
									A12	Connection Problem	10/03/2022	10/03/2022
							F2203	Imaging Required	A072201	High Impedance	10/03/2022	10/03/2022
									A12	Connection Problem	10/03/2022	10/03/2022
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A072201	High Impedance	10/03/2022	10/03/2022
									A12	Connection Problem	10/03/2022	10/03/2022
							F08	Hospitalization or Prolonged Hospitalization	A072201	High Impedance	10/03/2022	10/03/2022

									A12	Connection Problem	10/03/2022	10/03/2022
							F1905	Device Revision or Replacement	A072201	High Impedance	10/03/2022	10/03/2022
									A12	Connection Problem	10/03/2022	10/03/2022
							F2203	Imaging Required	A072201	High Impedance	10/03/2022	10/03/2022
									A12	Connection Problem	10/03/2022	10/03/2022
					CT2985	Active	E2403	No Clinical Signs, Symptoms or Conditions	A072201	High Impedance	10/03/2022	10/03/2022
									A12	Connection Problem	10/03/2022	10/03/2022
							F08	Hospitalization or Prolonged Hospitalization	A072201	High Impedance	10/03/2022	10/03/2022
									A12	Connection Problem	10/03/2022	10/03/2022
							F1905	Device Revision or Replacement	A072201	High Impedance	10/03/2022	10/03/2022
									A12	Connection Problem	10/03/2022	10/03/2022
							F2203	Imaging Required	A072201	High Impedance	10/03/2022	10/03/2022
									A12	Connection Problem	10/03/2022	10/03/2022
					CT3001	Implantable metal	E2403	No Clinical Signs, Symptoms or Conditions	A072201	High Impedance	10/03/2022	10/03/2022
									A12	Connection Problem	10/03/2022	10/03/2022
							F08	Hospitalization or Prolonged Hospitalization	A072201	High Impedance	10/03/2022	10/03/2022
									A12	Connection Problem	10/03/2022	10/03/2022
							F1905	Device Revision or Replacement	A072201	High Impedance	10/03/2022	10/03/2022
									A12	Connection Problem	10/03/2022	10/03/2022
							F2203	Imaging Required	A072201	High Impedance	10/03/2022	10/03/2022
									A12	Connection Problem	10/03/2022	10/03/2022
					CT3002	Implantable synthetic polymer	E2403	No Clinical Signs, Symptoms or Conditions	A072201	High Impedance	10/03/2022	10/03/2022
									A12	Connection Problem	10/03/2022	10/03/2022
							F08	Hospitalization or Prolonged Hospitalization	A072201	High Impedance	10/03/2022	10/03/2022
									A12	Connection Problem	10/03/2022	10/03/2022
							F1905	Device Revision or Replacement	A072201	High Impedance	10/03/2022	10/03/2022
									A12	Connection Problem	10/03/2022	10/03/2022
							F2203	Imaging Required	A072201	High Impedance	10/03/2022	10/03/2022
									A12	Connection Problem	10/03/2022	10/03/2022
					CT3006	Implantable inorganic material	E2403	No Clinical Signs, Symptoms or Conditions	A072201	High Impedance	10/03/2022	10/03/2022
									A12	Connection Problem	10/03/2022	10/03/2022
							F08	Hospitalization or Prolonged Hospitalization	A072201	High Impedance	10/03/2022	10/03/2022
									A12	Connection Problem	10/03/2022	10/03/2022
							F1905	Device Revision or Replacement	A072201	High Impedance	10/03/2022	10/03/2022
									A12	Connection Problem	10/03/2022	10/03/2022
							F2203	Imaging Required	A072201	High Impedance	10/03/2022	10/03/2022
									A12	Connection Problem	10/03/2022	10/03/2022
					CT314	Electrically powered	E2403	No Clinical Signs, Symptoms or Conditions	A072201	High Impedance	10/03/2022	10/03/2022
									A12	Connection Problem	10/03/2022	10/03/2022
							F08	Hospitalization or Prolonged Hospitalization	A072201	High Impedance	10/03/2022	10/03/2022
									A12	Connection Problem	10/03/2022	10/03/2022
							F1905	Device Revision or Replacement	A072201	High Impedance	10/03/2022	10/03/2022
									A12	Connection Problem	10/03/2022	10/03/2022
							F2203	Imaging Required	A072201	High Impedance	10/03/2022	10/03/2022
									A12	Connection Problem	10/03/2022	10/03/2022
					CT321	Long-term surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A072201	High Impedance	10/03/2022	10/03/2022
									A12	Connection Problem	10/03/2022	10/03/2022
							F08	Hospitalization or Prolonged Hospitalization	A072201	High Impedance	10/03/2022	10/03/2022
									A12	Connection Problem	10/03/2022	10/03/2022
							F1905	Device Revision or Replacement	A072201	High Impedance	10/03/2022	10/03/2022
									A12	Connection Problem	10/03/2022	10/03/2022
							F2203	Imaging Required	A072201	High Impedance	10/03/2022	10/03/2022
									A12	Connection Problem	10/03/2022	10/03/2022
					CT449	Generators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A072201	High Impedance	10/03/2022	10/03/2022
									A12	Connection Problem	10/03/2022	10/03/2022
							F08	Hospitalization or Prolonged Hospitalization	A072201	High Impedance	10/03/2022	10/03/2022
									A12	Connection Problem	10/03/2022	10/03/2022
							F1905	Device Revision or Replacement	A072201	High Impedance	10/03/2022	10/03/2022
									A12	Connection Problem	10/03/2022	10/03/2022
							F2203	Imaging Required	A072201	High Impedance	10/03/2022	10/03/2022
									A12	Connection Problem	10/03/2022	10/03/2022
					CT981	Single use	E2403	No Clinical Signs, Symptoms or Conditions	A072201	High Impedance	10/03/2022	10/03/2022
									A12	Connection Problem	10/03/2022	10/03/2022
							F08	Hospitalization or Prolonged Hospitalization	A072201	High Impedance	10/03/2022	10/03/2022
									A12	Connection Problem	10/03/2022	10/03/2022
							F1905	Device Revision or Replacement	A072201	High Impedance	10/03/2022	10/03/2022
									A12	Connection Problem	10/03/2022	10/03/2022
							F2203	Imaging Required	A072201	High Impedance	10/03/2022	10/03/2022
									A12	Connection Problem	10/03/2022	10/03/2022
					CT983	Surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A072201	High Impedance	10/03/2022	10/03/2022
									A12	Connection Problem	10/03/2022	10/03/2022
							F08	Hospitalization or Prolonged Hospitalization	A072201	High Impedance	10/03/2022	10/03/2022
									A12	Connection Problem	10/03/2022	10/03/2022
							F1905	Device Revision or Replacement	A072201	High Impedance	10/03/2022	10/03/2022
									A12	Connection Problem	10/03/2022	10/03/2022
							F2203	Imaging Required	A072201	High Impedance	10/03/2022	10/03/2022
									A12	Connection Problem	10/03/2022	10/03/2022
6177	2021	All other reportable incidents	35852	Single-chamber implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	27/07/2021	27/07/2021
							F1901	Additional Surgery	A070504	Premature Discharge of Battery	27/07/2021	27/07/2021
6178	2022	All other reportable incidents	37265	Dual-chamber implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F1901	Additional Surgery	A070504	Premature Discharge of Battery		
6179	2021	All other reportable incidents	37265	Dual-chamber implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A070910	Undersensing		
							F26	No Health Consequences or Impact	A070910	Undersensing		
6180	2021	All other reportable incidents	17882	Physiologic-monitoring defibrillation system	CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A1205	Loose or Intermittent Connection		
							F27	Problem Identified During Non-Clinical Procedure	A1205	Loose or Intermittent Connection		
6181	2022	All other reportable incidents	47265	Dual-chamber implantable pacemaker, rate-responsive	CT141	Electrodes/Leads and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A090805	Incorrect Measurement		
							F26	No Health Consequences or Impact	A090805	Incorrect Measurement		
					CT145	Cardiovascular devices	E2403	No Clinical Signs, Symptoms or Conditions	A090805	Incorrect Measurement		

							F26	No Health Consequences or Impact	A090805	Incorrect Measurement		
					CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A090805	Incorrect Measurement		
							F26	No Health Consequences or Impact	A090805	Incorrect Measurement		
					CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A090805	Incorrect Measurement		
							F26	No Health Consequences or Impact	A090805	Incorrect Measurement		
					CT233	Surgical	E2403	No Clinical Signs, Symptoms or Conditions	A090805	Incorrect Measurement		
							F26	No Health Consequences or Impact	A090805	Incorrect Measurement		
					CT2406	Implantable	E2403	No Clinical Signs, Symptoms or Conditions	A090805	Incorrect Measurement		
							F26	No Health Consequences or Impact	A090805	Incorrect Measurement		
					CT2407	Active implantable	E2403	No Clinical Signs, Symptoms or Conditions	A090805	Incorrect Measurement		
							F26	No Health Consequences or Impact	A090805	Incorrect Measurement		
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A090805	Incorrect Measurement		
							F26	No Health Consequences or Impact	A090805	Incorrect Measurement		
					CT2727	Active implant devices/ancillaries	E2403	No Clinical Signs, Symptoms or Conditions	A090805	Incorrect Measurement		
							F26	No Health Consequences or Impact	A090805	Incorrect Measurement		
					CT2750	Implantable cardiac electrical stimulators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A090805	Incorrect Measurement		
							F26	No Health Consequences or Impact	A090805	Incorrect Measurement		
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A090805	Incorrect Measurement		
							F26	No Health Consequences or Impact	A090805	Incorrect Measurement		
					CT2985	Active	E2403	No Clinical Signs, Symptoms or Conditions	A090805	Incorrect Measurement		
							F26	No Health Consequences or Impact	A090805	Incorrect Measurement		
					CT3001	Implantable metal	E2403	No Clinical Signs, Symptoms or Conditions	A090805	Incorrect Measurement		
							F26	No Health Consequences or Impact	A090805	Incorrect Measurement		
					CT3002	Implantable synthetic polymer	E2403	No Clinical Signs, Symptoms or Conditions	A090805	Incorrect Measurement		
							F26	No Health Consequences or Impact	A090805	Incorrect Measurement		
					CT3006	Implantable inorganic material	E2403	No Clinical Signs, Symptoms or Conditions	A090805	Incorrect Measurement		
							F26	No Health Consequences or Impact	A090805	Incorrect Measurement		
					CT314	Electrically powered	E2403	No Clinical Signs, Symptoms or Conditions	A090805	Incorrect Measurement		
							F26	No Health Consequences or Impact	A090805	Incorrect Measurement		
					CT321	Long-term surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A090805	Incorrect Measurement		
							F26	No Health Consequences or Impact	A090805	Incorrect Measurement		
					CT449	Generators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A090805	Incorrect Measurement		
							F26	No Health Consequences or Impact	A090805	Incorrect Measurement		
					CT981	Single use	E2403	No Clinical Signs, Symptoms or Conditions	A090805	Incorrect Measurement		
							F26	No Health Consequences or Impact	A090805	Incorrect Measurement		
					CT983	Surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A090805	Incorrect Measurement		
							F26	No Health Consequences or Impact	A090805	Incorrect Measurement		
6182	2022	All other reportable incidents	17882	Physiologic-monitoring defibrillation system	CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A071301	Failure to Deliver Shock/Stimulation		
							F27	Problem Identified During Non-Clinical Procedure	A071301	Failure to Deliver Shock/Stimulation		
6183	2022	All other reportable incidents	17882	Physiologic-monitoring defibrillation system	CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A07	Electrical /Electronic Property Problem		
							F27	Problem Identified During Non-Clinical Procedure	A07	Electrical /Electronic Property Problem		
6184	2022	All other reportable incidents	37265	Dual-chamber implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E0601	Arrhythmia	A0711	Interrogation Problem		
									A071201	Failure to Convert Rhythm		
									A0908	Incorrect, Inadequate or Imprecise Result or Readings		
							E2403	No Clinical Signs, Symptoms or Conditions	A0711	Interrogation Problem		
									A071201	Failure to Convert Rhythm		
									A0908	Incorrect, Inadequate or Imprecise Result or Readings		
							F23	Unexpected Medical Intervention	A0711	Interrogation Problem		
									A071201	Failure to Convert Rhythm		
									A0908	Incorrect, Inadequate or Imprecise Result or Readings		
6185	2021	All other reportable incidents	47270	Cardiac resynchronization therapy implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	04/08/2021	04/08/2021
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery	04/08/2021	04/08/2021
6186	2021	All other reportable incidents	35852	Single-chamber implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem	13/08/2021	13/08/2021
									A070504	Premature Discharge of Battery	13/08/2021	13/08/2021
							F08	Hospitalization or Prolonged Hospitalization	A0705	Battery Problem	13/08/2021	13/08/2021
									A070504	Premature Discharge of Battery	13/08/2021	13/08/2021
							F1903	Device Explantation	A0705	Battery Problem	13/08/2021	13/08/2021
									A070504	Premature Discharge of Battery	13/08/2021	13/08/2021
6187	2021	All other reportable incidents	36241	Pacemaker/defibrillator lead	CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070909	Oversensing		
									A0712	Pacing Problem		
									A072202	Low Impedance		
							F15	Recognized Device or Procedural Complication	A070909	Oversensing		
									A0712	Pacing Problem		
									A072202	Low Impedance		
6188	2021	All other reportable incidents	47265	Dual-chamber implantable pacemaker, rate-responsive	CT141	Electrodes/Leads and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070909	Oversensing		
									A0712	Pacing Problem		
									A072202	Low Impedance		
							F15	Recognized Device or Procedural Complication	A070909	Oversensing		
									A0712	Pacing Problem		
									A072202	Low Impedance		
					CT145	Cardiovascular devices	E2403	No Clinical Signs, Symptoms or Conditions	A070909	Oversensing		
									A0712	Pacing Problem		
									A072202	Low Impedance		
							F15	Recognized Device or Procedural Complication	A070909	Oversensing		
									A0712	Pacing Problem		
									A072202	Low Impedance		
					CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070909	Oversensing		
									A0712	Pacing Problem		
									A072202	Low Impedance		
							F15	Recognized Device or Procedural Complication	A070909	Oversensing		
									A0712	Pacing Problem		
									A072202	Low Impedance		
					CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A070909	Oversensing		
									A0712	Pacing Problem		
									A072202	Low Impedance		

		F15	Recognized Device or Procedural Complication	A070909	Oversensing			
				A0712	Pacing Problem			
				A072202	Low Impedance			
CT233	Surgical	E2403	No Clinical Signs, Symptoms or Conditions	A070909	Oversensing			
				A0712	Pacing Problem			
				A072202	Low Impedance			
		F15	Recognized Device or Procedural Complication	A070909	Oversensing			
				A0712	Pacing Problem			
				A072202	Low Impedance			
CT2406	Implantable	E2403	No Clinical Signs, Symptoms or Conditions	A070909	Oversensing			
				A0712	Pacing Problem			
				A072202	Low Impedance			
		F15	Recognized Device or Procedural Complication	A070909	Oversensing			
				A0712	Pacing Problem			
				A072202	Low Impedance			
CT2407	Active implantable	E2403	No Clinical Signs, Symptoms or Conditions	A070909	Oversensing			
				A0712	Pacing Problem			
				A072202	Low Impedance			
		F15	Recognized Device or Procedural Complication	A070909	Oversensing			
				A0712	Pacing Problem			
				A072202	Low Impedance			
CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A070909	Oversensing			
				A0712	Pacing Problem			
				A072202	Low Impedance			
		F15	Recognized Device or Procedural Complication	A070909	Oversensing			
				A0712	Pacing Problem			
				A072202	Low Impedance			
CT2727	Active implant devices/ancillaries	E2403	No Clinical Signs, Symptoms or Conditions	A070909	Oversensing			
				A0712	Pacing Problem			
				A072202	Low Impedance			
		F15	Recognized Device or Procedural Complication	A070909	Oversensing			
				A0712	Pacing Problem			
				A072202	Low Impedance			
CT2750	Implantable cardiac electrical stimulators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070909	Oversensing			
				A0712	Pacing Problem			
				A072202	Low Impedance			
		F15	Recognized Device or Procedural Complication	A070909	Oversensing			
				A0712	Pacing Problem			
				A072202	Low Impedance			
CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A070909	Oversensing			
				A0712	Pacing Problem			
				A072202	Low Impedance			
		F15	Recognized Device or Procedural Complication	A070909	Oversensing			
				A0712	Pacing Problem			
				A072202	Low Impedance			
CT2985	Active	E2403	No Clinical Signs, Symptoms or Conditions	A070909	Oversensing			
				A0712	Pacing Problem			
				A072202	Low Impedance			
		F15	Recognized Device or Procedural Complication	A070909	Oversensing			
				A0712	Pacing Problem			
				A072202	Low Impedance			
CT3001	Implantable metal	E2403	No Clinical Signs, Symptoms or Conditions	A070909	Oversensing			
				A0712	Pacing Problem			
				A072202	Low Impedance			
		F15	Recognized Device or Procedural Complication	A070909	Oversensing			
				A0712	Pacing Problem			
				A072202	Low Impedance			
CT3002	Implantable synthetic polymer	E2403	No Clinical Signs, Symptoms or Conditions	A070909	Oversensing			
				A0712	Pacing Problem			
				A072202	Low Impedance			
		F15	Recognized Device or Procedural Complication	A070909	Oversensing			
				A0712	Pacing Problem			
				A072202	Low Impedance			
CT3006	Implantable inorganic material	E2403	No Clinical Signs, Symptoms or Conditions	A070909	Oversensing			
				A0712	Pacing Problem			
				A072202	Low Impedance			
		F15	Recognized Device or Procedural Complication	A070909	Oversensing			
				A0712	Pacing Problem			
				A072202	Low Impedance			
CT314	Electrically powered	E2403	No Clinical Signs, Symptoms or Conditions	A070909	Oversensing			
				A0712	Pacing Problem			
				A072202	Low Impedance			
		F15	Recognized Device or Procedural Complication	A070909	Oversensing			
				A0712	Pacing Problem			
				A072202	Low Impedance			
CT321	Long-term surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A070909	Oversensing			
				A0712	Pacing Problem			
				A072202	Low Impedance			
		F15	Recognized Device or Procedural Complication	A070909	Oversensing			
				A0712	Pacing Problem			
				A072202	Low Impedance			
CT449	Generators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070909	Oversensing			
				A0712	Pacing Problem			
				A072202	Low Impedance			
		F15	Recognized Device or Procedural Complication	A070909	Oversensing			
				A0712	Pacing Problem			
				A072202	Low Impedance			
CT981	Single use	E2403	No Clinical Signs, Symptoms or Conditions	A070909	Oversensing			

								A0712	Pacing Problem		
								A072202	Low Impedance		
						F15	Recognized Device or Procedural Complication	A070909	Oversensing		
								A0712	Pacing Problem		
								A072202	Low Impedance		
					CT983	Surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A070909	Oversensing		
								A0712	Pacing Problem		
								A072202	Low Impedance		
							F15	Recognized Device or Procedural Complication	A070909	Oversensing		
									A0712	Pacing Problem		
									A072202	Low Impedance		
6189	2021	All other reportable incidents	47265	Dual-chamber implantable pacemaker, rate-responsive	CT141	Electrodes/Leads and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	30/08/2021	30/08/2021
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	30/08/2021	30/08/2021
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	30/08/2021	30/08/2021
					CT145	Cardiovascular devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	30/08/2021	30/08/2021
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	30/08/2021	30/08/2021
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	30/08/2021	30/08/2021
					CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	30/08/2021	30/08/2021
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	30/08/2021	30/08/2021
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	30/08/2021	30/08/2021
					CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	30/08/2021	30/08/2021
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	30/08/2021	30/08/2021
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	30/08/2021	30/08/2021
					CT233	Surgical	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	30/08/2021	30/08/2021
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	30/08/2021	30/08/2021
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	30/08/2021	30/08/2021
					CT2406	Implantable	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	30/08/2021	30/08/2021
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	30/08/2021	30/08/2021
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	30/08/2021	30/08/2021
					CT2407	Active implantable	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	30/08/2021	30/08/2021
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	30/08/2021	30/08/2021
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	30/08/2021	30/08/2021
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	30/08/2021	30/08/2021
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	30/08/2021	30/08/2021
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	30/08/2021	30/08/2021
					CT2727	Active implant devices/ancillaries	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	30/08/2021	30/08/2021
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	30/08/2021	30/08/2021
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	30/08/2021	30/08/2021
					CT2750	Implantable cardiac electrical stimulators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	30/08/2021	30/08/2021
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	30/08/2021	30/08/2021
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	30/08/2021	30/08/2021
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	30/08/2021	30/08/2021
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	30/08/2021	30/08/2021
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	30/08/2021	30/08/2021
					CT2985	Active	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	30/08/2021	30/08/2021
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	30/08/2021	30/08/2021
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	30/08/2021	30/08/2021
					CT3001	Implantable metal	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	30/08/2021	30/08/2021
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	30/08/2021	30/08/2021
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	30/08/2021	30/08/2021
					CT3002	Implantable synthetic polymer	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	30/08/2021	30/08/2021
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	30/08/2021	30/08/2021
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	30/08/2021	30/08/2021
					CT3006	Implantable inorganic material	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	30/08/2021	30/08/2021
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	30/08/2021	30/08/2021
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	30/08/2021	30/08/2021
					CT314	Electrically powered	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	30/08/2021	30/08/2021
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	30/08/2021	30/08/2021
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	30/08/2021	30/08/2021
					CT321	Long-term surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	30/08/2021	30/08/2021
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	30/08/2021	30/08/2021
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	30/08/2021	30/08/2021
					CT449	Generators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	30/08/2021	30/08/2021
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	30/08/2021	30/08/2021
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	30/08/2021	30/08/2021
					CT981	Single use	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	30/08/2021	30/08/2021
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	30/08/2021	30/08/2021
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	30/08/2021	30/08/2021
					CT983	Surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	30/08/2021	30/08/2021
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	30/08/2021	30/08/2021
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	30/08/2021	30/08/2021
6190	2021	All other reportable incidents	35852	Single-chamber implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E2104	Electric Shock	A070909	Oversensing		
									A071302	Inappropriate/Inadequate Shock/Stimulation		
									A090801	Signal Artifact/Noise		
							E2403	No Clinical Signs, Symptoms or Conditions	A070909	Oversensing		
									A071302	Inappropriate/Inadequate Shock/Stimulation		
									A090801	Signal Artifact/Noise		
							F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A070909	Oversensing		
									A071302	Inappropriate/Inadequate Shock/Stimulation		
									A090801	Signal Artifact/Noise		
6191	2021	All other reportable incidents	47265	Dual-chamber implantable pacemaker, rate-responsive	CT141	Electrodes/Leads and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem	24/09/2021	24/09/2021
									A070504	Premature Discharge of Battery	24/09/2021	24/09/2021
							F08	Hospitalization or Prolonged Hospitalization	A0705	Battery Problem	24/09/2021	24/09/2021
									A070504	Premature Discharge of Battery	24/09/2021	24/09/2021
							F1905	Device Revision or Replacement	A0705	Battery Problem	24/09/2021	24/09/2021
									A070504	Premature Discharge of Battery	24/09/2021	24/09/2021
					CT145	Cardiovascular devices	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem	24/09/2021	24/09/2021
									A070504	Premature Discharge of Battery	24/09/2021	24/09/2021

		F08	Hospitalization or Prolonged Hospitalization	A0705	Battery Problem	24/09/2021	24/09/2021
				A070504	Premature Discharge of Battery	24/09/2021	24/09/2021
		F1905	Device Revision or Replacement	A0705	Battery Problem	24/09/2021	24/09/2021
				A070504	Premature Discharge of Battery	24/09/2021	24/09/2021
CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem	24/09/2021	24/09/2021
				A070504	Premature Discharge of Battery	24/09/2021	24/09/2021
		F08	Hospitalization or Prolonged Hospitalization	A0705	Battery Problem	24/09/2021	24/09/2021
				A070504	Premature Discharge of Battery	24/09/2021	24/09/2021
		F1905	Device Revision or Replacement	A0705	Battery Problem	24/09/2021	24/09/2021
				A070504	Premature Discharge of Battery	24/09/2021	24/09/2021
CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem	24/09/2021	24/09/2021
				A070504	Premature Discharge of Battery	24/09/2021	24/09/2021
		F08	Hospitalization or Prolonged Hospitalization	A0705	Battery Problem	24/09/2021	24/09/2021
				A070504	Premature Discharge of Battery	24/09/2021	24/09/2021
		F1905	Device Revision or Replacement	A0705	Battery Problem	24/09/2021	24/09/2021
				A070504	Premature Discharge of Battery	24/09/2021	24/09/2021
CT233	Surgical	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem	24/09/2021	24/09/2021
				A070504	Premature Discharge of Battery	24/09/2021	24/09/2021
		F08	Hospitalization or Prolonged Hospitalization	A0705	Battery Problem	24/09/2021	24/09/2021
				A070504	Premature Discharge of Battery	24/09/2021	24/09/2021
		F1905	Device Revision or Replacement	A0705	Battery Problem	24/09/2021	24/09/2021
				A070504	Premature Discharge of Battery	24/09/2021	24/09/2021
CT2406	Implantable	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem	24/09/2021	24/09/2021
				A070504	Premature Discharge of Battery	24/09/2021	24/09/2021
		F08	Hospitalization or Prolonged Hospitalization	A0705	Battery Problem	24/09/2021	24/09/2021
				A070504	Premature Discharge of Battery	24/09/2021	24/09/2021
		F1905	Device Revision or Replacement	A0705	Battery Problem	24/09/2021	24/09/2021
				A070504	Premature Discharge of Battery	24/09/2021	24/09/2021
CT2407	Active implantable	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem	24/09/2021	24/09/2021
				A070504	Premature Discharge of Battery	24/09/2021	24/09/2021
		F08	Hospitalization or Prolonged Hospitalization	A0705	Battery Problem	24/09/2021	24/09/2021
				A070504	Premature Discharge of Battery	24/09/2021	24/09/2021
		F1905	Device Revision or Replacement	A0705	Battery Problem	24/09/2021	24/09/2021
				A070504	Premature Discharge of Battery	24/09/2021	24/09/2021
CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem	24/09/2021	24/09/2021
				A070504	Premature Discharge of Battery	24/09/2021	24/09/2021
		F08	Hospitalization or Prolonged Hospitalization	A0705	Battery Problem	24/09/2021	24/09/2021
				A070504	Premature Discharge of Battery	24/09/2021	24/09/2021
		F1905	Device Revision or Replacement	A0705	Battery Problem	24/09/2021	24/09/2021
				A070504	Premature Discharge of Battery	24/09/2021	24/09/2021
CT2727	Active implant devices/ancillaries	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem	24/09/2021	24/09/2021
				A070504	Premature Discharge of Battery	24/09/2021	24/09/2021
		F08	Hospitalization or Prolonged Hospitalization	A0705	Battery Problem	24/09/2021	24/09/2021
				A070504	Premature Discharge of Battery	24/09/2021	24/09/2021
		F1905	Device Revision or Replacement	A0705	Battery Problem	24/09/2021	24/09/2021
				A070504	Premature Discharge of Battery	24/09/2021	24/09/2021
CT2750	Implantable cardiac electrical stimulators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem	24/09/2021	24/09/2021
				A070504	Premature Discharge of Battery	24/09/2021	24/09/2021
		F08	Hospitalization or Prolonged Hospitalization	A0705	Battery Problem	24/09/2021	24/09/2021
				A070504	Premature Discharge of Battery	24/09/2021	24/09/2021
		F1905	Device Revision or Replacement	A0705	Battery Problem	24/09/2021	24/09/2021
				A070504	Premature Discharge of Battery	24/09/2021	24/09/2021
CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem	24/09/2021	24/09/2021
				A070504	Premature Discharge of Battery	24/09/2021	24/09/2021
		F08	Hospitalization or Prolonged Hospitalization	A0705	Battery Problem	24/09/2021	24/09/2021
				A070504	Premature Discharge of Battery	24/09/2021	24/09/2021
		F1905	Device Revision or Replacement	A0705	Battery Problem	24/09/2021	24/09/2021
				A070504	Premature Discharge of Battery	24/09/2021	24/09/2021
CT2985	Active	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem	24/09/2021	24/09/2021
				A070504	Premature Discharge of Battery	24/09/2021	24/09/2021
		F08	Hospitalization or Prolonged Hospitalization	A0705	Battery Problem	24/09/2021	24/09/2021
				A070504	Premature Discharge of Battery	24/09/2021	24/09/2021
		F1905	Device Revision or Replacement	A0705	Battery Problem	24/09/2021	24/09/2021
				A070504	Premature Discharge of Battery	24/09/2021	24/09/2021
CT3001	Implantable metal	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem	24/09/2021	24/09/2021
				A070504	Premature Discharge of Battery	24/09/2021	24/09/2021
		F08	Hospitalization or Prolonged Hospitalization	A0705	Battery Problem	24/09/2021	24/09/2021
				A070504	Premature Discharge of Battery	24/09/2021	24/09/2021
		F1905	Device Revision or Replacement	A0705	Battery Problem	24/09/2021	24/09/2021
				A070504	Premature Discharge of Battery	24/09/2021	24/09/2021
CT3002	Implantable synthetic polymer	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem	24/09/2021	24/09/2021
				A070504	Premature Discharge of Battery	24/09/2021	24/09/2021
		F08	Hospitalization or Prolonged Hospitalization	A0705	Battery Problem	24/09/2021	24/09/2021
				A070504	Premature Discharge of Battery	24/09/2021	24/09/2021
		F1905	Device Revision or Replacement	A0705	Battery Problem	24/09/2021	24/09/2021
				A070504	Premature Discharge of Battery	24/09/2021	24/09/2021
CT3006	Implantable inorganic material	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem	24/09/2021	24/09/2021
				A070504	Premature Discharge of Battery	24/09/2021	24/09/2021
		F08	Hospitalization or Prolonged Hospitalization	A0705	Battery Problem	24/09/2021	24/09/2021
				A070504	Premature Discharge of Battery	24/09/2021	24/09/2021
		F1905	Device Revision or Replacement	A0705	Battery Problem	24/09/2021	24/09/2021
				A070504	Premature Discharge of Battery	24/09/2021	24/09/2021
CT314	Electrically powered	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem	24/09/2021	24/09/2021
				A070504	Premature Discharge of Battery	24/09/2021	24/09/2021
		F08	Hospitalization or Prolonged Hospitalization	A0705	Battery Problem	24/09/2021	24/09/2021
				A070504	Premature Discharge of Battery	24/09/2021	24/09/2021
		F1905	Device Revision or Replacement	A0705	Battery Problem	24/09/2021	24/09/2021
				A070504	Premature Discharge of Battery	24/09/2021	24/09/2021

					CT321	Long-term surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem	24/09/2021	24/09/2021
									A070504	Premature Discharge of Battery	24/09/2021	24/09/2021
							F08	Hospitalization or Prolonged Hospitalization	A0705	Battery Problem	24/09/2021	24/09/2021
									A070504	Premature Discharge of Battery	24/09/2021	24/09/2021
							F1905	Device Revision or Replacement	A0705	Battery Problem	24/09/2021	24/09/2021
									A070504	Premature Discharge of Battery	24/09/2021	24/09/2021
					CT449	Generators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem	24/09/2021	24/09/2021
									A070504	Premature Discharge of Battery	24/09/2021	24/09/2021
							F08	Hospitalization or Prolonged Hospitalization	A0705	Battery Problem	24/09/2021	24/09/2021
									A070504	Premature Discharge of Battery	24/09/2021	24/09/2021
							F1905	Device Revision or Replacement	A0705	Battery Problem	24/09/2021	24/09/2021
									A070504	Premature Discharge of Battery	24/09/2021	24/09/2021
					CT981	Single use	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem	24/09/2021	24/09/2021
									A070504	Premature Discharge of Battery	24/09/2021	24/09/2021
							F08	Hospitalization or Prolonged Hospitalization	A0705	Battery Problem	24/09/2021	24/09/2021
									A070504	Premature Discharge of Battery	24/09/2021	24/09/2021
							F1905	Device Revision or Replacement	A0705	Battery Problem	24/09/2021	24/09/2021
									A070504	Premature Discharge of Battery	24/09/2021	24/09/2021
					CT983	Surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem	24/09/2021	24/09/2021
									A070504	Premature Discharge of Battery	24/09/2021	24/09/2021
							F08	Hospitalization or Prolonged Hospitalization	A0705	Battery Problem	24/09/2021	24/09/2021
									A070504	Premature Discharge of Battery	24/09/2021	24/09/2021
							F1905	Device Revision or Replacement	A0705	Battery Problem	24/09/2021	24/09/2021
									A070504	Premature Discharge of Battery	24/09/2021	24/09/2021
6192	2021	All other reportable incidents	35852	Single-chamber implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A070909	Oversensing		
							F26	No Health Consequences or Impact	A070909	Oversensing		
6193	2021	All other reportable incidents	35852	Single-chamber implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem	23/08/2021	23/08/2021
									A070504	Premature Discharge of Battery	23/08/2021	23/08/2021
							F08	Hospitalization or Prolonged Hospitalization	A0705	Battery Problem	23/08/2021	23/08/2021
									A070504	Premature Discharge of Battery	23/08/2021	23/08/2021
							F1903	Device Explantation	A0705	Battery Problem	23/08/2021	23/08/2021
									A070504	Premature Discharge of Battery	23/08/2021	23/08/2021
6194	2021	All other reportable incidents	37265	Dual-chamber implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E060102	Atrial Fibrillation	A070910	Undersensing		
									A0712	Pacing Problem		
							E2403	No Clinical Signs, Symptoms or Conditions	A070910	Undersensing		
									A0712	Pacing Problem		
							F26	No Health Consequences or Impact	A070910	Undersensing		
									A0712	Pacing Problem		
6195	2021	All other reportable incidents	47263	Cardiac resynchronization therapy implantable pacemaker	CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A072201	High Impedance		
							F26	No Health Consequences or Impact	A072201	High Impedance		
					CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A072201	High Impedance		
							F26	No Health Consequences or Impact	A072201	High Impedance		
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A072201	High Impedance		
							F26	No Health Consequences or Impact	A072201	High Impedance		
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A072201	High Impedance		
							F26	No Health Consequences or Impact	A072201	High Impedance		
6196	2021	All other reportable incidents	36241	Pacemaker/defibrillator lead	CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A072201	High Impedance		
							F2203	Imaging Required	A072201	High Impedance		
6197	2021	All other reportable incidents	47265	Dual-chamber implantable pacemaker, rate-responsive	CT141	Electrodes/Leads and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070909	Oversensing		
									A072201	High Impedance		
									A090801	Signal Artifact/Noise		
							F26	No Health Consequences or Impact	A070909	Oversensing		
									A072201	High Impedance		
									A090801	Signal Artifact/Noise		
					CT145	Cardiovascular devices	E2403	No Clinical Signs, Symptoms or Conditions	A070909	Oversensing		
									A072201	High Impedance		
									A090801	Signal Artifact/Noise		
							F26	No Health Consequences or Impact	A070909	Oversensing		
									A072201	High Impedance		
									A090801	Signal Artifact/Noise		
					CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070909	Oversensing		
									A072201	High Impedance		
									A090801	Signal Artifact/Noise		
							F26	No Health Consequences or Impact	A070909	Oversensing		
									A072201	High Impedance		
									A090801	Signal Artifact/Noise		
					CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A070909	Oversensing		
									A072201	High Impedance		
									A090801	Signal Artifact/Noise		
							F26	No Health Consequences or Impact	A070909	Oversensing		
									A072201	High Impedance		
									A090801	Signal Artifact/Noise		
					CT233	Surgical	E2403	No Clinical Signs, Symptoms or Conditions	A070909	Oversensing		
									A072201	High Impedance		
									A090801	Signal Artifact/Noise		
							F26	No Health Consequences or Impact	A070909	Oversensing		
									A072201	High Impedance		
									A090801	Signal Artifact/Noise		
					CT2406	Implantable	E2403	No Clinical Signs, Symptoms or Conditions	A070909	Oversensing		
									A072201	High Impedance		
									A090801	Signal Artifact/Noise		
							F26	No Health Consequences or Impact	A070909	Oversensing		
									A072201	High Impedance		
									A090801	Signal Artifact/Noise		
					CT2407	Active implantable	E2403	No Clinical Signs, Symptoms or Conditions	A070909	Oversensing		
									A072201	High Impedance		
									A090801	Signal Artifact/Noise		
							F26	No Health Consequences or Impact	A070909	Oversensing		

									A072201	High Impedance		
									A090801	Signal Artifact/Noise		
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A070909	Oversensing		
									A072201	High Impedance		
									A090801	Signal Artifact/Noise		
							F26	No Health Consequences or Impact	A070909	Oversensing		
									A072201	High Impedance		
									A090801	Signal Artifact/Noise		
					CT2727	Active implant devices/ancillaries	E2403	No Clinical Signs, Symptoms or Conditions	A070909	Oversensing		
									A072201	High Impedance		
									A090801	Signal Artifact/Noise		
							F26	No Health Consequences or Impact	A070909	Oversensing		
									A072201	High Impedance		
									A090801	Signal Artifact/Noise		
					CT2750	Implantable cardiac electrical stimulators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070909	Oversensing		
									A072201	High Impedance		
									A090801	Signal Artifact/Noise		
							F26	No Health Consequences or Impact	A070909	Oversensing		
									A072201	High Impedance		
									A090801	Signal Artifact/Noise		
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A070909	Oversensing		
									A072201	High Impedance		
									A090801	Signal Artifact/Noise		
							F26	No Health Consequences or Impact	A070909	Oversensing		
									A072201	High Impedance		
									A090801	Signal Artifact/Noise		
					CT2985	Active	E2403	No Clinical Signs, Symptoms or Conditions	A070909	Oversensing		
									A072201	High Impedance		
									A090801	Signal Artifact/Noise		
							F26	No Health Consequences or Impact	A070909	Oversensing		
									A072201	High Impedance		
									A090801	Signal Artifact/Noise		
					CT3001	Implantable metal	E2403	No Clinical Signs, Symptoms or Conditions	A070909	Oversensing		
									A072201	High Impedance		
									A090801	Signal Artifact/Noise		
							F26	No Health Consequences or Impact	A070909	Oversensing		
									A072201	High Impedance		
									A090801	Signal Artifact/Noise		
					CT3002	Implantable synthetic polymer	E2403	No Clinical Signs, Symptoms or Conditions	A070909	Oversensing		
									A072201	High Impedance		
									A090801	Signal Artifact/Noise		
							F26	No Health Consequences or Impact	A070909	Oversensing		
									A072201	High Impedance		
									A090801	Signal Artifact/Noise		
					CT3006	Implantable inorganic material	E2403	No Clinical Signs, Symptoms or Conditions	A070909	Oversensing		
									A072201	High Impedance		
									A090801	Signal Artifact/Noise		
							F26	No Health Consequences or Impact	A070909	Oversensing		
									A072201	High Impedance		
									A090801	Signal Artifact/Noise		
					CT314	Electrically powered	E2403	No Clinical Signs, Symptoms or Conditions	A070909	Oversensing		
									A072201	High Impedance		
									A090801	Signal Artifact/Noise		
							F26	No Health Consequences or Impact	A070909	Oversensing		
									A072201	High Impedance		
									A090801	Signal Artifact/Noise		
					CT321	Long-term surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A070909	Oversensing		
									A072201	High Impedance		
									A090801	Signal Artifact/Noise		
							F26	No Health Consequences or Impact	A070909	Oversensing		
									A072201	High Impedance		
									A090801	Signal Artifact/Noise		
					CT449	Generators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070909	Oversensing		
									A072201	High Impedance		
									A090801	Signal Artifact/Noise		
							F26	No Health Consequences or Impact	A070909	Oversensing		
									A072201	High Impedance		
									A090801	Signal Artifact/Noise		
					CT981	Single use	E2403	No Clinical Signs, Symptoms or Conditions	A070909	Oversensing		
									A072201	High Impedance		
									A090801	Signal Artifact/Noise		
							F26	No Health Consequences or Impact	A070909	Oversensing		
									A072201	High Impedance		
									A090801	Signal Artifact/Noise		
					CT983	Surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A070909	Oversensing		
									A072201	High Impedance		
									A090801	Signal Artifact/Noise		
							F26	No Health Consequences or Impact	A070909	Oversensing		
									A072201	High Impedance		
									A090801	Signal Artifact/Noise		
6198	2021	All other reportable incidents	47265	Dual-chamber implantable pacemaker, rate-responsive	CT141	Electrodes/Leads and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT145	Cardiovascular devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		

					CT233	Surgical	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT2406	Implantable	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT2407	Active implantable	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT2727	Active implant devices/ancillaries	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT2750	Implantable cardiac electrical stimulators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT2985	Active	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT3001	Implantable metal	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT3002	Implantable synthetic polymer	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT3006	Implantable inorganic material	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT314	Electrically powered	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT321	Long-term surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT449	Generators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT981	Single use	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT983	Surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
6199	2022	All other reportable incidents	47270	Cardiac resynchronization therapy implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A070909	Oversensing		
							F23	Unexpected Medical Intervention	A070909	Oversensing		
6200	2021	All other reportable incidents	47263	Cardiac resynchronization therapy implantable pacemaker	CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070909	Oversensing		
									A090801	Signal Artifact/Noise		
							F13	Misdiagnosis/ Misclassification	A070909	Oversensing		
									A090801	Signal Artifact/Noise		
					CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A070909	Oversensing		
									A090801	Signal Artifact/Noise		
							F13	Misdiagnosis/ Misclassification	A070909	Oversensing		
									A090801	Signal Artifact/Noise		
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A070909	Oversensing		
									A090801	Signal Artifact/Noise		
							F13	Misdiagnosis/ Misclassification	A070909	Oversensing		
									A090801	Signal Artifact/Noise		
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A070909	Oversensing		
									A090801	Signal Artifact/Noise		
							F13	Misdiagnosis/ Misclassification	A070909	Oversensing		
									A090801	Signal Artifact/Noise		
6201	2022	All other reportable incidents	35223	Endocardial/interventricular septal pacing lead	CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A040101	Fracture		
									A070909	Oversensing		
									A090801	Signal Artifact/Noise		
							F26	No Health Consequences or Impact	A040101	Fracture		
									A070909	Oversensing		
									A090801	Signal Artifact/Noise		
6202	2022	All other reportable incidents	36241	Pacemaker/defibrillator lead	CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070909	Oversensing		
									A1301	Failure to Read Input Signal		
							F26	No Health Consequences or Impact	A070909	Oversensing		
									A1301	Failure to Read Input Signal		
6203	2022	All other reportable incidents	35223	Endocardial/interventricular septal pacing lead	CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A040101	Fracture		
									A0701	Capturing Problem		
									A070909	Oversensing		
									A0722	Impedance Problem		
									A090801	Signal Artifact/Noise		
							F26	No Health Consequences or Impact	A040101	Fracture		
									A0701	Capturing Problem		
									A070909	Oversensing		
									A0722	Impedance Problem		
									A090801	Signal Artifact/Noise		
6204	2022	All other reportable incidents	36241	Pacemaker/defibrillator lead	CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A051201	Device Dislodged or Dislocated		
									A070101	Failure to Capture		
							F08	Hospitalization or Prolonged Hospitalization	A051201	Device Dislodged or Dislocated		
									A070101	Failure to Capture		
							F1905	Device Revision or Replacement	A051201	Device Dislodged or Dislocated		
									A070101	Failure to Capture		
6205	2022	All other reportable incidents	37265	Dual-chamber implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
6206	2021	All other reportable incidents	47265	Dual-chamber implantable pacemaker, rate-responsive	CT141	Electrodes/Leads and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT145	Cardiovascular devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		

					CT233	Surgical	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT2406	Implantable	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT2407	Active implantable	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT2727	Active implant devices/ancillaries	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT2750	Implantable cardiac electrical stimulators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT2985	Active	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT3001	Implantable metal	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT3002	Implantable synthetic polymer	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT3006	Implantable inorganic material	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT314	Electrically powered	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT321	Long-term surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT449	Generators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT981	Single use	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT983	Surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
6207	2022	All other reportable incidents	35852	Single-chamber implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem	11/08/2022	11/08/2022
									A070504	Premature Discharge of Battery	11/08/2022	11/08/2022
							F08	Hospitalization or Prolonged Hospitalization	A0705	Battery Problem	11/08/2022	11/08/2022
									A070504	Premature Discharge of Battery	11/08/2022	11/08/2022
							F1905	Device Revision or Replacement	A0705	Battery Problem	11/08/2022	11/08/2022
									A070504	Premature Discharge of Battery	11/08/2022	11/08/2022
6208	2022	All other reportable incidents	35223	Endocardial/interventricular septal pacing lead	CT1684	Cardiac pacemakers and associated devices	E0506	Hemorrhage/Blood Loss/Bleeding	A040101	Fracture		
							E060101	Asystole	A040101	Fracture		
							E0605	Cardiac Tamponade	A040101	Fracture		
							E0619	Pericardial Effusion	A040101	Fracture		
							E1906	Unspecified Infection	A040101	Fracture		
							E2321	Low Blood Pressure/ Hypotension	A040101	Fracture		
							E2403	No Clinical Signs, Symptoms or Conditions	A040101	Fracture		
							F1203	Life Threatening Illness or Injury	A040101	Fracture		
							F1903	Device Explantation	A040101	Fracture		
							F2203	Imaging Required	A040101	Fracture		
							F2303	Medication Required	A040101	Fracture		
6209	2021	All other reportable incidents	35223	Endocardial/interventricular septal pacing lead	CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070909	Oversensing		
							F26	No Health Consequences or Impact	A070909	Oversensing		
6210	2021	All other reportable incidents	60789	Intracardiac pacemaker	CT130	Catheters and associated devices	E011903	Syncope/Fainting	A070910	Undersensing		
							E2403	No Clinical Signs, Symptoms or Conditions	A070910	Undersensing		
							F1203	Life Threatening Illness or Injury	A070910	Undersensing		
							F23	Unexpected Medical Intervention	A070910	Undersensing		
					CT145	Cardiovascular devices	E011903	Syncope/Fainting	A070910	Undersensing		
							E2403	No Clinical Signs, Symptoms or Conditions	A070910	Undersensing		
							F1203	Life Threatening Illness or Injury	A070910	Undersensing		
							F23	Unexpected Medical Intervention	A070910	Undersensing		
					CT1684	Cardiac pacemakers and associated devices	E011903	Syncope/Fainting	A070910	Undersensing		
							E2403	No Clinical Signs, Symptoms or Conditions	A070910	Undersensing		
							F1203	Life Threatening Illness or Injury	A070910	Undersensing		
							F23	Unexpected Medical Intervention	A070910	Undersensing		
					CT1685	Cardiac pacemakers	E011903	Syncope/Fainting	A070910	Undersensing		
							E2403	No Clinical Signs, Symptoms or Conditions	A070910	Undersensing		
							F1203	Life Threatening Illness or Injury	A070910	Undersensing		
							F23	Unexpected Medical Intervention	A070910	Undersensing		
					CT233	Surgical	E011903	Syncope/Fainting	A070910	Undersensing		
							E2403	No Clinical Signs, Symptoms or Conditions	A070910	Undersensing		
							F1203	Life Threatening Illness or Injury	A070910	Undersensing		
							F23	Unexpected Medical Intervention	A070910	Undersensing		
					CT2406	Implantable	E011903	Syncope/Fainting	A070910	Undersensing		
							E2403	No Clinical Signs, Symptoms or Conditions	A070910	Undersensing		
							F1203	Life Threatening Illness or Injury	A070910	Undersensing		
							F23	Unexpected Medical Intervention	A070910	Undersensing		
					CT2407	Active implantable	E011903	Syncope/Fainting	A070910	Undersensing		
							E2403	No Clinical Signs, Symptoms or Conditions	A070910	Undersensing		
							F1203	Life Threatening Illness or Injury	A070910	Undersensing		
							F23	Unexpected Medical Intervention	A070910	Undersensing		
					CT257	Implantable cardiac pacemakers	E011903	Syncope/Fainting	A070910	Undersensing		
							E2403	No Clinical Signs, Symptoms or Conditions	A070910	Undersensing		
							F1203	Life Threatening Illness or Injury	A070910	Undersensing		
							F23	Unexpected Medical Intervention	A070910	Undersensing		
					CT2727	Active implant devices/ancillaries	E011903	Syncope/Fainting	A070910	Undersensing		
							E2403	No Clinical Signs, Symptoms or Conditions	A070910	Undersensing		
							F1203	Life Threatening Illness or Injury	A070910	Undersensing		
							F23	Unexpected Medical Intervention	A070910	Undersensing		
					CT2750	Implantable cardiac electrical stimulators and associated devices	E011903	Syncope/Fainting	A070910	Undersensing		

							E2403	No Clinical Signs, Symptoms or Conditions	A070910	Undersensing		
							F1203	Life Threatening Illness or Injury	A070910	Undersensing		
							F23	Unexpected Medical Intervention	A070910	Undersensing		
					CT2818	Catheterization/Cannulation/Fluid path devices	E011903	Syncope/Fainting	A070910	Undersensing		
							E2403	No Clinical Signs, Symptoms or Conditions	A070910	Undersensing		
							F1203	Life Threatening Illness or Injury	A070910	Undersensing		
							F23	Unexpected Medical Intervention	A070910	Undersensing		
					CT2866	Implantable cardiac electrical stimulators	E011903	Syncope/Fainting	A070910	Undersensing		
							E2403	No Clinical Signs, Symptoms or Conditions	A070910	Undersensing		
							F1203	Life Threatening Illness or Injury	A070910	Undersensing		
							F23	Unexpected Medical Intervention	A070910	Undersensing		
					CT2985	Active	E011903	Syncope/Fainting	A070910	Undersensing		
							E2403	No Clinical Signs, Symptoms or Conditions	A070910	Undersensing		
							F1203	Life Threatening Illness or Injury	A070910	Undersensing		
							F23	Unexpected Medical Intervention	A070910	Undersensing		
					CT3001	Implantable metal	E011903	Syncope/Fainting	A070910	Undersensing		
							E2403	No Clinical Signs, Symptoms or Conditions	A070910	Undersensing		
							F1203	Life Threatening Illness or Injury	A070910	Undersensing		
							F23	Unexpected Medical Intervention	A070910	Undersensing		
					CT3002	Implantable synthetic polymer	E011903	Syncope/Fainting	A070910	Undersensing		
							E2403	No Clinical Signs, Symptoms or Conditions	A070910	Undersensing		
							F1203	Life Threatening Illness or Injury	A070910	Undersensing		
							F23	Unexpected Medical Intervention	A070910	Undersensing		
					CT3006	Implantable inorganic material	E011903	Syncope/Fainting	A070910	Undersensing		
							E2403	No Clinical Signs, Symptoms or Conditions	A070910	Undersensing		
							F1203	Life Threatening Illness or Injury	A070910	Undersensing		
							F23	Unexpected Medical Intervention	A070910	Undersensing		
					CT314	Electrically powered	E011903	Syncope/Fainting	A070910	Undersensing		
							E2403	No Clinical Signs, Symptoms or Conditions	A070910	Undersensing		
							F1203	Life Threatening Illness or Injury	A070910	Undersensing		
							F23	Unexpected Medical Intervention	A070910	Undersensing		
					CT321	Long-term surgical invasive	E011903	Syncope/Fainting	A070910	Undersensing		
							E2403	No Clinical Signs, Symptoms or Conditions	A070910	Undersensing		
							F1203	Life Threatening Illness or Injury	A070910	Undersensing		
							F23	Unexpected Medical Intervention	A070910	Undersensing		
					CT981	Single use	E011903	Syncope/Fainting	A070910	Undersensing		
							E2403	No Clinical Signs, Symptoms or Conditions	A070910	Undersensing		
							F1203	Life Threatening Illness or Injury	A070910	Undersensing		
							F23	Unexpected Medical Intervention	A070910	Undersensing		
					CT983	Surgical invasive	E011903	Syncope/Fainting	A070910	Undersensing		
							E2403	No Clinical Signs, Symptoms or Conditions	A070910	Undersensing		
							F1203	Life Threatening Illness or Injury	A070910	Undersensing		
							F23	Unexpected Medical Intervention	A070910	Undersensing		
6211	2021	All other reportable incidents	37265	Dual-chamber implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A070909	Oversensing	28/04/2021	28/04/2021
							F26	No Health Consequences or Impact	A070909	Oversensing	28/04/2021	28/04/2021
6212	2021	All other reportable incidents	36241	Pacemaker/defibrillator lead	CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A0203	Defective Device		
									A072202	Low Impedance		
							F08	Hospitalization or Prolonged Hospitalization	A0203	Defective Device		
									A072202	Low Impedance		
							F1905	Device Revision or Replacement	A0203	Defective Device		
									A072202	Low Impedance		
6213	2021	All other reportable incidents	35852	Single-chamber implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem	10/09/2021	10/09/2021
									A070504	Premature Discharge of Battery	10/09/2021	10/09/2021
							F08	Hospitalization or Prolonged Hospitalization	A0705	Battery Problem	10/09/2021	10/09/2021
									A070504	Premature Discharge of Battery	10/09/2021	10/09/2021
							F1905	Device Revision or Replacement	A0705	Battery Problem	10/09/2021	10/09/2021
									A070504	Premature Discharge of Battery	10/09/2021	10/09/2021
6214	2021	All other reportable incidents	35852	Single-chamber implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A070910	Undersensing	27/07/2021	27/07/2021
									A071201	Failure to Convert Rhythm	27/07/2021	27/07/2021
									A0713	Defibrillation/Stimulation Problem	27/07/2021	27/07/2021
									A072201	High Impedance	27/07/2021	27/07/2021
							F05	Delay to Treatment/ Therapy	A070910	Undersensing	27/07/2021	27/07/2021
									A071201	Failure to Convert Rhythm	27/07/2021	27/07/2021
									A0713	Defibrillation/Stimulation Problem	27/07/2021	27/07/2021
									A072201	High Impedance	27/07/2021	27/07/2021
							F08	Hospitalization or Prolonged Hospitalization	A070910	Undersensing	27/07/2021	27/07/2021
									A071201	Failure to Convert Rhythm	27/07/2021	27/07/2021
									A0713	Defibrillation/Stimulation Problem	27/07/2021	27/07/2021
									A072201	High Impedance	27/07/2021	27/07/2021
							F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A070910	Undersensing	27/07/2021	27/07/2021
									A071201	Failure to Convert Rhythm	27/07/2021	27/07/2021
									A0713	Defibrillation/Stimulation Problem	27/07/2021	27/07/2021
									A072201	High Impedance	27/07/2021	27/07/2021
							F1905	Device Revision or Replacement	A070910	Undersensing	27/07/2021	27/07/2021
									A071201	Failure to Convert Rhythm	27/07/2021	27/07/2021
									A0713	Defibrillation/Stimulation Problem	27/07/2021	27/07/2021
									A072201	High Impedance	27/07/2021	27/07/2021
							F2203	Imaging Required	A070910	Undersensing	27/07/2021	27/07/2021
									A071201	Failure to Convert Rhythm	27/07/2021	27/07/2021
									A0713	Defibrillation/Stimulation Problem	27/07/2021	27/07/2021
									A072201	High Impedance	27/07/2021	27/07/2021
							F2306	Resuscitation	A070910	Undersensing	27/07/2021	27/07/2021
									A071201	Failure to Convert Rhythm	27/07/2021	27/07/2021
									A0713	Defibrillation/Stimulation Problem	27/07/2021	27/07/2021
									A072201	High Impedance	27/07/2021	27/07/2021
6215	2021	All other reportable incidents	47265	Dual-chamber implantable pacemaker, rate-responsive	CT141	Electrodes/Leads and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	27/07/2021	27/07/2021
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	27/07/2021	27/07/2021
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	27/07/2021	27/07/2021

					CT145	Cardiovascular devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	27/07/2021	27/07/2021
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	27/07/2021	27/07/2021
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	27/07/2021	27/07/2021
					CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	27/07/2021	27/07/2021
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	27/07/2021	27/07/2021
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	27/07/2021	27/07/2021
					CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	27/07/2021	27/07/2021
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	27/07/2021	27/07/2021
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	27/07/2021	27/07/2021
					CT233	Surgical	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	27/07/2021	27/07/2021
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	27/07/2021	27/07/2021
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	27/07/2021	27/07/2021
					CT2406	Implantable	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	27/07/2021	27/07/2021
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	27/07/2021	27/07/2021
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	27/07/2021	27/07/2021
					CT2407	Active implantable	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	27/07/2021	27/07/2021
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	27/07/2021	27/07/2021
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	27/07/2021	27/07/2021
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	27/07/2021	27/07/2021
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	27/07/2021	27/07/2021
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	27/07/2021	27/07/2021
					CT2727	Active implant devices/ancillaries	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	27/07/2021	27/07/2021
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	27/07/2021	27/07/2021
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	27/07/2021	27/07/2021
					CT2750	Implantable cardiac electrical stimulators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	27/07/2021	27/07/2021
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	27/07/2021	27/07/2021
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	27/07/2021	27/07/2021
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	27/07/2021	27/07/2021
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	27/07/2021	27/07/2021
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	27/07/2021	27/07/2021
					CT2985	Active	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	27/07/2021	27/07/2021
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	27/07/2021	27/07/2021
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	27/07/2021	27/07/2021
					CT3001	Implantable metal	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	27/07/2021	27/07/2021
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	27/07/2021	27/07/2021
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	27/07/2021	27/07/2021
					CT3002	Implantable synthetic polymer	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	27/07/2021	27/07/2021
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	27/07/2021	27/07/2021
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	27/07/2021	27/07/2021
					CT3006	Implantable inorganic material	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	27/07/2021	27/07/2021
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	27/07/2021	27/07/2021
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	27/07/2021	27/07/2021
					CT314	Electrically powered	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	27/07/2021	27/07/2021
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	27/07/2021	27/07/2021
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	27/07/2021	27/07/2021
					CT321	Long-term surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	27/07/2021	27/07/2021
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	27/07/2021	27/07/2021
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	27/07/2021	27/07/2021
					CT449	Generators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	27/07/2021	27/07/2021
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	27/07/2021	27/07/2021
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	27/07/2021	27/07/2021
					CT981	Single use	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	27/07/2021	27/07/2021
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	27/07/2021	27/07/2021
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	27/07/2021	27/07/2021
					CT983	Surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	27/07/2021	27/07/2021
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	27/07/2021	27/07/2021
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	27/07/2021	27/07/2021
6216	2022	All other reportable incidents	47265	Dual-chamber implantable pacemaker, rate-responsive	CT141	Electrodes/Leads and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	30/03/2022	30/03/2022
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery	30/03/2022	30/03/2022
					CT145	Cardiovascular devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	30/03/2022	30/03/2022
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery	30/03/2022	30/03/2022
					CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	30/03/2022	30/03/2022
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery	30/03/2022	30/03/2022
					CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	30/03/2022	30/03/2022
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery	30/03/2022	30/03/2022
					CT233	Surgical	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	30/03/2022	30/03/2022
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery	30/03/2022	30/03/2022
					CT2406	Implantable	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	30/03/2022	30/03/2022
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery	30/03/2022	30/03/2022
					CT2407	Active implantable	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	30/03/2022	30/03/2022
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery	30/03/2022	30/03/2022
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	30/03/2022	30/03/2022
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery	30/03/2022	30/03/2022
					CT2727	Active implant devices/ancillaries	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	30/03/2022	30/03/2022
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery	30/03/2022	30/03/2022
					CT2750	Implantable cardiac electrical stimulators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	30/03/2022	30/03/2022
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery	30/03/2022	30/03/2022
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	30/03/2022	30/03/2022
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery	30/03/2022	30/03/2022
					CT2985	Active	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	30/03/2022	30/03/2022
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery	30/03/2022	30/03/2022
					CT3001	Implantable metal	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	30/03/2022	30/03/2022
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery	30/03/2022	30/03/2022
					CT3002	Implantable synthetic polymer	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	30/03/2022	30/03/2022
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery	30/03/2022	30/03/2022
					CT3006	Implantable inorganic material	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	30/03/2022	30/03/2022
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery	30/03/2022	30/03/2022
					CT314	Electrically powered	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	30/03/2022	30/03/2022

						F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery	30/03/2022	30/03/2022
					CT321	Long-term surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	30/03/2022	30/03/2022
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery	30/03/2022	30/03/2022
					CT449	Generators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	30/03/2022	30/03/2022
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery	30/03/2022	30/03/2022
					CT981	Single use	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	30/03/2022	30/03/2022
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery	30/03/2022	30/03/2022
					CT983	Surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	30/03/2022	30/03/2022
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery	30/03/2022	30/03/2022
6217	2022	All other reportable incidents	35223	Endocardial/interventricular septal pacing lead	CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A0401	Break		
									A070909	Oversensing		
									A090801	Signal Artifact/Noise		
							F12	Serious Injury/ Illness/ Impairment	A0401	Break		
									A070909	Oversensing		
									A090801	Signal Artifact/Noise		
							F23	Unexpected Medical Intervention	A0401	Break		
									A070909	Oversensing		
									A090801	Signal Artifact/Noise		
							F26	No Health Consequences or Impact	A0401	Break		
									A070909	Oversensing		
									A090801	Signal Artifact/Noise		
6218	2022	All other reportable incidents	17882	Physiologic-monitoring defibrillation system	CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070803	Failure to Power Up		
							F27	Problem Identified During Non-Clinical Procedure	A070803	Failure to Power Up		
6219	2021	All other reportable incidents	35852	Single-chamber implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
6220	2022	All other reportable incidents	47267	Single-chamber implantable pacemaker, rate-responsive	CT141	Electrodes/Leads and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A13	Communication or Transmission Problem		
							F26	No Health Consequences or Impact	A13	Communication or Transmission Problem		
					CT145	Cardiovascular devices	E2403	No Clinical Signs, Symptoms or Conditions	A13	Communication or Transmission Problem		
							F26	No Health Consequences or Impact	A13	Communication or Transmission Problem		
					CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A13	Communication or Transmission Problem		
							F26	No Health Consequences or Impact	A13	Communication or Transmission Problem		
					CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A13	Communication or Transmission Problem		
							F26	No Health Consequences or Impact	A13	Communication or Transmission Problem		
					CT233	Surgical	E2403	No Clinical Signs, Symptoms or Conditions	A13	Communication or Transmission Problem		
							F26	No Health Consequences or Impact	A13	Communication or Transmission Problem		
					CT2406	Implantable	E2403	No Clinical Signs, Symptoms or Conditions	A13	Communication or Transmission Problem		
							F26	No Health Consequences or Impact	A13	Communication or Transmission Problem		
					CT2407	Active implantable	E2403	No Clinical Signs, Symptoms or Conditions	A13	Communication or Transmission Problem		
							F26	No Health Consequences or Impact	A13	Communication or Transmission Problem		
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A13	Communication or Transmission Problem		
							F26	No Health Consequences or Impact	A13	Communication or Transmission Problem		
					CT2727	Active implant devices/ancillaries	E2403	No Clinical Signs, Symptoms or Conditions	A13	Communication or Transmission Problem		
							F26	No Health Consequences or Impact	A13	Communication or Transmission Problem		
					CT2750	Implantable cardiac electrical stimulators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A13	Communication or Transmission Problem		
							F26	No Health Consequences or Impact	A13	Communication or Transmission Problem		
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A13	Communication or Transmission Problem		
							F26	No Health Consequences or Impact	A13	Communication or Transmission Problem		
					CT2985	Active	E2403	No Clinical Signs, Symptoms or Conditions	A13	Communication or Transmission Problem		
							F26	No Health Consequences or Impact	A13	Communication or Transmission Problem		
					CT3001	Implantable metal	E2403	No Clinical Signs, Symptoms or Conditions	A13	Communication or Transmission Problem		
							F26	No Health Consequences or Impact	A13	Communication or Transmission Problem		
					CT3002	Implantable synthetic polymer	E2403	No Clinical Signs, Symptoms or Conditions	A13	Communication or Transmission Problem		
							F26	No Health Consequences or Impact	A13	Communication or Transmission Problem		
					CT3006	Implantable inorganic material	E2403	No Clinical Signs, Symptoms or Conditions	A13	Communication or Transmission Problem		
							F26	No Health Consequences or Impact	A13	Communication or Transmission Problem		
					CT314	Electrically powered	E2403	No Clinical Signs, Symptoms or Conditions	A13	Communication or Transmission Problem		
							F26	No Health Consequences or Impact	A13	Communication or Transmission Problem		
					CT321	Long-term surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A13	Communication or Transmission Problem		
							F26	No Health Consequences or Impact	A13	Communication or Transmission Problem		
					CT449	Generators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A13	Communication or Transmission Problem		
							F26	No Health Consequences or Impact	A13	Communication or Transmission Problem		
					CT981	Single use	E2403	No Clinical Signs, Symptoms or Conditions	A13	Communication or Transmission Problem		
							F26	No Health Consequences or Impact	A13	Communication or Transmission Problem		
					CT983	Surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A13	Communication or Transmission Problem		
							F26	No Health Consequences or Impact	A13	Communication or Transmission Problem		
6221	2021	All other reportable incidents	17882	Physiologic-monitoring defibrillation system	CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070803	Failure to Power Up		
							F27	Problem Identified During Non-Clinical Procedure	A070803	Failure to Power Up		
6222	2021	All other reportable incidents	17882	Physiologic-monitoring defibrillation system	CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070803	Failure to Power Up		
							F27	Problem Identified During Non-Clinical Procedure	A070803	Failure to Power Up		
6223	2021	All other reportable incidents	17882	Physiologic-monitoring defibrillation system	CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070803	Failure to Power Up		
							F27	Problem Identified During Non-Clinical Procedure	A070803	Failure to Power Up		
6224	2021	All other reportable incidents	17882	Physiologic-monitoring defibrillation system	CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070803	Failure to Power Up		
							F27	Problem Identified During Non-Clinical Procedure	A070803	Failure to Power Up		
6225	2022	All other reportable incidents	35852	Single-chamber implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E020201	Anxiety	A0705	Battery Problem	03/05/2022	03/05/2022
									A070504	Premature Discharge of Battery	03/05/2022	03/05/2022
							E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem	03/05/2022	03/05/2022
									A070504	Premature Discharge of Battery	03/05/2022	03/05/2022
							F08	Hospitalization or Prolonged Hospitalization	A0705	Battery Problem	03/05/2022	03/05/2022
									A070504	Premature Discharge of Battery	03/05/2022	03/05/2022
							F1905	Device Revision or Replacement	A0705	Battery Problem	03/05/2022	03/05/2022
									A070504	Premature Discharge of Battery	03/05/2022	03/05/2022
6226	2022	All other reportable incidents	63188	Left-ventricular wireless pacing system sensor/transmitter	CT141	Electrodes/Leads and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A26	Insufficient Device Problem Information	03/03/2022	03/03/2022
					CT145	Cardiovascular devices	E2403	No Clinical Signs, Symptoms or Conditions	A26	Insufficient Device Problem Information	03/03/2022	03/03/2022
					CT1577	Wireless monitoring system transmitters	E2403	No Clinical Signs, Symptoms or Conditions	A26	Insufficient Device Problem Information	03/03/2022	03/03/2022
					CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A26	Insufficient Device Problem Information	03/03/2022	03/03/2022
					CT2317	Transmitters/Receivers/Transceivers	E2403	No Clinical Signs, Symptoms or Conditions	A26	Insufficient Device Problem Information	03/03/2022	03/03/2022
					CT233	Surgical	E2403	No Clinical Signs, Symptoms or Conditions	A26	Insufficient Device Problem Information	03/03/2022	03/03/2022
					CT2406	Implantable	E2403	No Clinical Signs, Symptoms or Conditions	A26	Insufficient Device Problem Information	03/03/2022	03/03/2022

					CT2407	Active implantable	E2403	No Clinical Signs, Symptoms or Conditions	A26	Insufficient Device Problem Information	03/03/2022	03/03/2022
					CT255	Transducers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A26	Insufficient Device Problem Information	03/03/2022	03/03/2022
					CT2726	Physiological monitoring/measuring devices	E2403	No Clinical Signs, Symptoms or Conditions	A26	Insufficient Device Problem Information	03/03/2022	03/03/2022
					CT2727	Active implant devices/ancillaries	E2403	No Clinical Signs, Symptoms or Conditions	A26	Insufficient Device Problem Information	03/03/2022	03/03/2022
					CT2750	Implantable cardiac electrical stimulators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A26	Insufficient Device Problem Information	03/03/2022	03/03/2022
					CT2885	Physiological monitoring/measuring system transmitters/receivers/transceivers	E2403	No Clinical Signs, Symptoms or Conditions	A26	Insufficient Device Problem Information	03/03/2022	03/03/2022
					CT2985	Active	E2403	No Clinical Signs, Symptoms or Conditions	A26	Insufficient Device Problem Information	03/03/2022	03/03/2022
					CT3001	Implantable metal	E2403	No Clinical Signs, Symptoms or Conditions	A26	Insufficient Device Problem Information	03/03/2022	03/03/2022
					CT3002	Implantable synthetic polymer	E2403	No Clinical Signs, Symptoms or Conditions	A26	Insufficient Device Problem Information	03/03/2022	03/03/2022
					CT3006	Implantable inorganic material	E2403	No Clinical Signs, Symptoms or Conditions	A26	Insufficient Device Problem Information	03/03/2022	03/03/2022
					CT314	Electrically powered	E2403	No Clinical Signs, Symptoms or Conditions	A26	Insufficient Device Problem Information	03/03/2022	03/03/2022
					CT321	Long-term surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A26	Insufficient Device Problem Information	03/03/2022	03/03/2022
					CT384	Sensors and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A26	Insufficient Device Problem Information	03/03/2022	03/03/2022
					CT449	Generators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A26	Insufficient Device Problem Information	03/03/2022	03/03/2022
					CT981	Single use	E2403	No Clinical Signs, Symptoms or Conditions	A26	Insufficient Device Problem Information	03/03/2022	03/03/2022
					CT983	Surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A26	Insufficient Device Problem Information	03/03/2022	03/03/2022
					CT994	Sensors	E2403	No Clinical Signs, Symptoms or Conditions	A26	Insufficient Device Problem Information	03/03/2022	03/03/2022
6227	2021	All other reportable incidents	35223	Endocardial/interventricular septal pacing lead	CT1684	Cardiac pacemakers and associated devices	E0505	Hematoma	A070909	Oversensing		
									A090801	Signal Artifact/Noise		
							E2403	No Clinical Signs, Symptoms or Conditions	A070909	Oversensing		
									A090801	Signal Artifact/Noise		
							F11	Non-Serious Injury/ Illness/ Impairment	A070909	Oversensing		
									A090801	Signal Artifact/Noise		
6228	2022	All other reportable incidents	35852	Single-chamber implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A072201	High Impedance		
							F26	No Health Consequences or Impact	A072201	High Impedance		
6229	2021	All other reportable incidents	17882	Physiologic-monitoring defibrillation system	CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A1205	Loose or Intermittent Connection		
							F27	Problem Identified During Non-Clinical Procedure	A1205	Loose or Intermittent Connection		
6230	2022	Death	17882	Physiologic-monitoring defibrillation system	CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A071204	No Pacing		
							F26	No Health Consequences or Impact	A071204	No Pacing		
6231	2022	All other reportable incidents	47265	Dual-chamber implantable pacemaker, rate-responsive	CT141	Electrodes/Leads and associated devices	E011903	Syncope/Fainting	A24	No Apparent Device Problem		
							E060109	Tachycardia	A24	No Apparent Device Problem		
							E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem		
							F1203	Life Threatening Illness or Injury	A24	No Apparent Device Problem		
							F23	Unexpected Medical Intervention	A24	No Apparent Device Problem		
					CT145	Cardiovascular devices	E011903	Syncope/Fainting	A24	No Apparent Device Problem		
							E060109	Tachycardia	A24	No Apparent Device Problem		
							E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem		
							F1203	Life Threatening Illness or Injury	A24	No Apparent Device Problem		
							F23	Unexpected Medical Intervention	A24	No Apparent Device Problem		
					CT1684	Cardiac pacemakers and associated devices	E011903	Syncope/Fainting	A24	No Apparent Device Problem		
							E060109	Tachycardia	A24	No Apparent Device Problem		
							E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem		
							F1203	Life Threatening Illness or Injury	A24	No Apparent Device Problem		
							F23	Unexpected Medical Intervention	A24	No Apparent Device Problem		
					CT1685	Cardiac pacemakers	E011903	Syncope/Fainting	A24	No Apparent Device Problem		
							E060109	Tachycardia	A24	No Apparent Device Problem		
							E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem		
							F1203	Life Threatening Illness or Injury	A24	No Apparent Device Problem		
							F23	Unexpected Medical Intervention	A24	No Apparent Device Problem		
					CT233	Surgical	E011903	Syncope/Fainting	A24	No Apparent Device Problem		
							E060109	Tachycardia	A24	No Apparent Device Problem		
							E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem		
							F1203	Life Threatening Illness or Injury	A24	No Apparent Device Problem		
							F23	Unexpected Medical Intervention	A24	No Apparent Device Problem		
					CT2406	Implantable	E011903	Syncope/Fainting	A24	No Apparent Device Problem		
							E060109	Tachycardia	A24	No Apparent Device Problem		
							E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem		
							F1203	Life Threatening Illness or Injury	A24	No Apparent Device Problem		
							F23	Unexpected Medical Intervention	A24	No Apparent Device Problem		
					CT2407	Active implantable	E011903	Syncope/Fainting	A24	No Apparent Device Problem		
							E060109	Tachycardia	A24	No Apparent Device Problem		
							E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem		
							F1203	Life Threatening Illness or Injury	A24	No Apparent Device Problem		
							F23	Unexpected Medical Intervention	A24	No Apparent Device Problem		
					CT257	Implantable cardiac pacemakers	E011903	Syncope/Fainting	A24	No Apparent Device Problem		
							E060109	Tachycardia	A24	No Apparent Device Problem		
							E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem		
							F1203	Life Threatening Illness or Injury	A24	No Apparent Device Problem		
							F23	Unexpected Medical Intervention	A24	No Apparent Device Problem		
					CT2727	Active implant devices/ancillaries	E011903	Syncope/Fainting	A24	No Apparent Device Problem		
							E060109	Tachycardia	A24	No Apparent Device Problem		
							E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem		
							F1203	Life Threatening Illness or Injury	A24	No Apparent Device Problem		
							F23	Unexpected Medical Intervention	A24	No Apparent Device Problem		
					CT2750	Implantable cardiac electrical stimulators and associated devices	E011903	Syncope/Fainting	A24	No Apparent Device Problem		
							E060109	Tachycardia	A24	No Apparent Device Problem		
							E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem		
							F1203	Life Threatening Illness or Injury	A24	No Apparent Device Problem		
							F23	Unexpected Medical Intervention	A24	No Apparent Device Problem		
					CT2866	Implantable cardiac electrical stimulators	E011903	Syncope/Fainting	A24	No Apparent Device Problem		
							E060109	Tachycardia	A24	No Apparent Device Problem		
							E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem		
							F1203	Life Threatening Illness or Injury	A24	No Apparent Device Problem		
							F23	Unexpected Medical Intervention	A24	No Apparent Device Problem		
					CT2985	Active	E011903	Syncope/Fainting	A24	No Apparent Device Problem		
							E060109	Tachycardia	A24	No Apparent Device Problem		

							E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem		
							F1203	Life Threatening Illness or Injury	A24	No Apparent Device Problem		
							F23	Unexpected Medical Intervention	A24	No Apparent Device Problem		
					CT3001	Implantable metal	E011903	Syncope/Fainting	A24	No Apparent Device Problem		
							E060109	Tachycardia	A24	No Apparent Device Problem		
							E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem		
							F1203	Life Threatening Illness or Injury	A24	No Apparent Device Problem		
							F23	Unexpected Medical Intervention	A24	No Apparent Device Problem		
					CT3002	Implantable synthetic polymer	E011903	Syncope/Fainting	A24	No Apparent Device Problem		
							E060109	Tachycardia	A24	No Apparent Device Problem		
							E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem		
							F1203	Life Threatening Illness or Injury	A24	No Apparent Device Problem		
							F23	Unexpected Medical Intervention	A24	No Apparent Device Problem		
					CT3006	Implantable inorganic material	E011903	Syncope/Fainting	A24	No Apparent Device Problem		
							E060109	Tachycardia	A24	No Apparent Device Problem		
							E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem		
							F1203	Life Threatening Illness or Injury	A24	No Apparent Device Problem		
							F23	Unexpected Medical Intervention	A24	No Apparent Device Problem		
					CT314	Electrically powered	E011903	Syncope/Fainting	A24	No Apparent Device Problem		
							E060109	Tachycardia	A24	No Apparent Device Problem		
							E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem		
							F1203	Life Threatening Illness or Injury	A24	No Apparent Device Problem		
							F23	Unexpected Medical Intervention	A24	No Apparent Device Problem		
					CT321	Long-term surgical invasive	E011903	Syncope/Fainting	A24	No Apparent Device Problem		
							E060109	Tachycardia	A24	No Apparent Device Problem		
							E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem		
							F1203	Life Threatening Illness or Injury	A24	No Apparent Device Problem		
							F23	Unexpected Medical Intervention	A24	No Apparent Device Problem		
					CT449	Generators and associated devices	E011903	Syncope/Fainting	A24	No Apparent Device Problem		
							E060109	Tachycardia	A24	No Apparent Device Problem		
							E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem		
							F1203	Life Threatening Illness or Injury	A24	No Apparent Device Problem		
							F23	Unexpected Medical Intervention	A24	No Apparent Device Problem		
					CT981	Single use	E011903	Syncope/Fainting	A24	No Apparent Device Problem		
							E060109	Tachycardia	A24	No Apparent Device Problem		
							E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem		
							F1203	Life Threatening Illness or Injury	A24	No Apparent Device Problem		
							F23	Unexpected Medical Intervention	A24	No Apparent Device Problem		
					CT983	Surgical invasive	E011903	Syncope/Fainting	A24	No Apparent Device Problem		
							E060109	Tachycardia	A24	No Apparent Device Problem		
							E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem		
							F1203	Life Threatening Illness or Injury	A24	No Apparent Device Problem		
							F23	Unexpected Medical Intervention	A24	No Apparent Device Problem		
6232	2022	All other reportable incidents	35223	Endocardial/interventricular septal pacing lead	CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A07	Electrical /Electronic Property Problem		
									A072202	Low Impedance		
							F12	Serious Injury/ Illness/ Impairment	A07	Electrical /Electronic Property Problem		
									A072202	Low Impedance		
							F23	Unexpected Medical Intervention	A07	Electrical /Electronic Property Problem		
									A072202	Low Impedance		
							F26	No Health Consequences or Impact	A07	Electrical /Electronic Property Problem		
									A072202	Low Impedance		
6233	2022	All other reportable incidents	47265	Dual-chamber implantable pacemaker, rate-responsive	CT141	Electrodes/Leads and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
					CT145	Cardiovascular devices	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
					CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
					CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
					CT233	Surgical	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
					CT2406	Implantable	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
					CT2407	Active implantable	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
					CT2727	Active implant devices/ancillaries	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
					CT2750	Implantable cardiac electrical stimulators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem		

									A070504	Premature Discharge of Battery			
							F26	No Health Consequences or Impact	A0705	Battery Problem			
									A070504	Premature Discharge of Battery			
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem			
									A070504	Premature Discharge of Battery			
							F26	No Health Consequences or Impact	A0705	Battery Problem			
									A070504	Premature Discharge of Battery			
					CT2985	Active	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem			
									A070504	Premature Discharge of Battery			
							F26	No Health Consequences or Impact	A0705	Battery Problem			
									A070504	Premature Discharge of Battery			
					CT3001	Implantable metal	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem			
									A070504	Premature Discharge of Battery			
							F26	No Health Consequences or Impact	A0705	Battery Problem			
									A070504	Premature Discharge of Battery			
					CT3002	Implantable synthetic polymer	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem			
									A070504	Premature Discharge of Battery			
							F26	No Health Consequences or Impact	A0705	Battery Problem			
									A070504	Premature Discharge of Battery			
					CT3006	Implantable inorganic material	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem			
									A070504	Premature Discharge of Battery			
							F26	No Health Consequences or Impact	A0705	Battery Problem			
									A070504	Premature Discharge of Battery			
					CT314	Electrically powered	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem			
									A070504	Premature Discharge of Battery			
							F26	No Health Consequences or Impact	A0705	Battery Problem			
									A070504	Premature Discharge of Battery			
					CT321	Long-term surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem			
									A070504	Premature Discharge of Battery			
							F26	No Health Consequences or Impact	A0705	Battery Problem			
									A070504	Premature Discharge of Battery			
					CT449	Generators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem			
									A070504	Premature Discharge of Battery			
							F26	No Health Consequences or Impact	A0705	Battery Problem			
									A070504	Premature Discharge of Battery			
					CT981	Single use	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem			
									A070504	Premature Discharge of Battery			
							F26	No Health Consequences or Impact	A0705	Battery Problem			
									A070504	Premature Discharge of Battery			
					CT983	Surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem			
									A070504	Premature Discharge of Battery			
							F26	No Health Consequences or Impact	A0705	Battery Problem			
									A070504	Premature Discharge of Battery			
6234	2021	All other reportable incidents	17882	Physiologic-monitoring defibrillation system	CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070803	Failure to Power Up			
							F27	Problem Identified During Non-Clinical Procedure	A070803	Failure to Power Up			
6235	2021	All other reportable incidents	36241	Pacemaker/defibrillator lead	CT1684	Cardiac pacemakers and associated devices	E2104	Electric Shock	A070909	Oversensing	13/10/2021	13/10/2021	
									A071302	Inappropriate/Inadequate Shock/Stimulation	13/10/2021	13/10/2021	
							E2403	No Clinical Signs, Symptoms or Conditions	A070909	Oversensing	13/10/2021	13/10/2021	
									A071302	Inappropriate/Inadequate Shock/Stimulation	13/10/2021	13/10/2021	
							F08	Hospitalization or Prolonged Hospitalization	A070909	Oversensing	13/10/2021	13/10/2021	
									A071302	Inappropriate/Inadequate Shock/Stimulation	13/10/2021	13/10/2021	
							F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A070909	Oversensing	13/10/2021	13/10/2021	
									A071302	Inappropriate/Inadequate Shock/Stimulation	13/10/2021	13/10/2021	
							F1905	Device Revision or Replacement	A070909	Oversensing	13/10/2021	13/10/2021	
									A071302	Inappropriate/Inadequate Shock/Stimulation	13/10/2021	13/10/2021	
							F2203	Imaging Required	A070909	Oversensing	13/10/2021	13/10/2021	
									A071302	Inappropriate/Inadequate Shock/Stimulation	13/10/2021	13/10/2021	
							F23	Unexpected Medical Intervention	A070909	Oversensing	13/10/2021	13/10/2021	
									A071302	Inappropriate/Inadequate Shock/Stimulation	13/10/2021	13/10/2021	
6236	2021	All other reportable incidents	35852	Single-chamber implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E2104	Electric Shock	A070909	Oversensing	13/10/2021	13/10/2021	
									A071302	Inappropriate/Inadequate Shock/Stimulation	13/10/2021	13/10/2021	
							E2403	No Clinical Signs, Symptoms or Conditions	A070909	Oversensing	13/10/2021	13/10/2021	
									A071302	Inappropriate/Inadequate Shock/Stimulation	13/10/2021	13/10/2021	
							F08	Hospitalization or Prolonged Hospitalization	A070909	Oversensing	13/10/2021	13/10/2021	
									A071302	Inappropriate/Inadequate Shock/Stimulation	13/10/2021	13/10/2021	
							F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A070909	Oversensing	13/10/2021	13/10/2021	
									A071302	Inappropriate/Inadequate Shock/Stimulation	13/10/2021	13/10/2021	
							F1905	Device Revision or Replacement	A070909	Oversensing	13/10/2021	13/10/2021	
									A071302	Inappropriate/Inadequate Shock/Stimulation	13/10/2021	13/10/2021	
							F2203	Imaging Required	A070909	Oversensing	13/10/2021	13/10/2021	
									A071302	Inappropriate/Inadequate Shock/Stimulation	13/10/2021	13/10/2021	
							F23	Unexpected Medical Intervention	A070909	Oversensing	13/10/2021	13/10/2021	
									A071302	Inappropriate/Inadequate Shock/Stimulation	13/10/2021	13/10/2021	
6237	2022	All other reportable incidents	36241	Pacemaker/defibrillator lead	CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A072201	High Impedance			
							F26	No Health Consequences or Impact	A072201	High Impedance			
6238	2021	All other reportable incidents	35223	Endocardial/interventricular septal pacing lead	CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A040101	Fracture	29/07/2021	29/07/2021	
							F26	No Health Consequences or Impact	A040101	Fracture	29/07/2021	29/07/2021	
6239	2021	All other reportable incidents	47270	Cardiac resynchronization therapy implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery			
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery			
6240	2022	All other reportable incidents	47267	Single-chamber implantable pacemaker, rate-responsive	CT141	Electrodes/Leads and associated devices	E0611	Heart Failure/Congestive Heart Failure	A070504	Premature Discharge of Battery			
							E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery			
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery			
					CT145	Cardiovascular devices	E0611	Heart Failure/Congestive Heart Failure	A070504	Premature Discharge of Battery			
							E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery			
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery			
					CT1684	Cardiac pacemakers and associated devices	E0611	Heart Failure/Congestive Heart Failure	A070504	Premature Discharge of Battery			
							E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery			
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery			

					CT1685	Cardiac pacemakers	E0611	Heart Failure/Congestive Heart Failure	A070504	Premature Discharge of Battery		
							E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT233	Surgical	E0611	Heart Failure/Congestive Heart Failure	A070504	Premature Discharge of Battery		
							E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT2406	Implantable	E0611	Heart Failure/Congestive Heart Failure	A070504	Premature Discharge of Battery		
							E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT2407	Active implantable	E0611	Heart Failure/Congestive Heart Failure	A070504	Premature Discharge of Battery		
							E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT257	Implantable cardiac pacemakers	E0611	Heart Failure/Congestive Heart Failure	A070504	Premature Discharge of Battery		
							E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT2727	Active implant devices/ancillaries	E0611	Heart Failure/Congestive Heart Failure	A070504	Premature Discharge of Battery		
							E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT2750	Implantable cardiac electrical stimulators and associated devices	E0611	Heart Failure/Congestive Heart Failure	A070504	Premature Discharge of Battery		
							E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT2866	Implantable cardiac electrical stimulators	E0611	Heart Failure/Congestive Heart Failure	A070504	Premature Discharge of Battery		
							E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT2985	Active	E0611	Heart Failure/Congestive Heart Failure	A070504	Premature Discharge of Battery		
							E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT3001	Implantable metal	E0611	Heart Failure/Congestive Heart Failure	A070504	Premature Discharge of Battery		
							E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT3002	Implantable synthetic polymer	E0611	Heart Failure/Congestive Heart Failure	A070504	Premature Discharge of Battery		
							E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT3006	Implantable inorganic material	E0611	Heart Failure/Congestive Heart Failure	A070504	Premature Discharge of Battery		
							E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT314	Electrically powered	E0611	Heart Failure/Congestive Heart Failure	A070504	Premature Discharge of Battery		
							E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT321	Long-term surgical invasive	E0611	Heart Failure/Congestive Heart Failure	A070504	Premature Discharge of Battery		
							E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT449	Generators and associated devices	E0611	Heart Failure/Congestive Heart Failure	A070504	Premature Discharge of Battery		
							E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT981	Single use	E0611	Heart Failure/Congestive Heart Failure	A070504	Premature Discharge of Battery		
							E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT983	Surgical invasive	E0611	Heart Failure/Congestive Heart Failure	A070504	Premature Discharge of Battery		
							E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
6241	2022	All other reportable incidents	35223	Endocardial/interventricular septal pacing lead	CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070909	Oversensing		
							F26	No Health Consequences or Impact	A070909	Oversensing		
6242	2022	All other reportable incidents	35223	Endocardial/interventricular septal pacing lead	CT1684	Cardiac pacemakers and associated devices	E011903	Syncope/Fainting	A070910	Undersensing		
							E2403	No Clinical Signs, Symptoms or Conditions	A070910	Undersensing		
							F08	Hospitalization or Prolonged Hospitalization	A070910	Undersensing		
							F1203	Life Threatening Illness or Injury	A070910	Undersensing		
6243	2021	All other reportable incidents	17882	Physiologic-monitoring defibrillation system	CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A071204	No Pacing		
							F27	Problem Identified During Non-Clinical Procedure	A071204	No Pacing		
6244	2021	All other reportable incidents	35852	Single-chamber implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem	25/08/2021	25/08/2021
									A070504	Premature Discharge of Battery	25/08/2021	25/08/2021
							F08	Hospitalization or Prolonged Hospitalization	A0705	Battery Problem	25/08/2021	25/08/2021
									A070504	Premature Discharge of Battery	25/08/2021	25/08/2021
							F1905	Device Revision or Replacement	A0705	Battery Problem	25/08/2021	25/08/2021
									A070504	Premature Discharge of Battery	25/08/2021	25/08/2021
6245	2021	All other reportable incidents	17882	Physiologic-monitoring defibrillation system	CT1684	Cardiac pacemakers and associated devices	E0602	Cardiac Arrest	A071301	Failure to Deliver Shock/Stimulation		
							E2403	No Clinical Signs, Symptoms or Conditions	A071301	Failure to Deliver Shock/Stimulation		
							F26	No Health Consequences or Impact	A071301	Failure to Deliver Shock/Stimulation		
6246	2021	All other reportable incidents	35852	Single-chamber implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem	15/09/2021	15/09/2021
									A070504	Premature Discharge of Battery	15/09/2021	15/09/2021
							F08	Hospitalization or Prolonged Hospitalization	A0705	Battery Problem	15/09/2021	15/09/2021
									A070504	Premature Discharge of Battery	15/09/2021	15/09/2021
							F1905	Device Revision or Replacement	A0705	Battery Problem	15/09/2021	15/09/2021
									A070504	Premature Discharge of Battery	15/09/2021	15/09/2021
6247	2021	All other reportable incidents	47265	Dual-chamber implantable pacemaker, rate-responsive	CT141	Electrodes/Leads and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	21/07/2021	21/07/2021
									A070909	Oversensing	21/07/2021	21/07/2021
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	21/07/2021	21/07/2021
									A070909	Oversensing	21/07/2021	21/07/2021
							F1903	Device Explantation	A070504	Premature Discharge of Battery	21/07/2021	21/07/2021
									A070909	Oversensing	21/07/2021	21/07/2021
					CT145	Cardiovascular devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	21/07/2021	21/07/2021
									A070909	Oversensing	21/07/2021	21/07/2021
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	21/07/2021	21/07/2021
									A070909	Oversensing	21/07/2021	21/07/2021
							F1903	Device Explantation	A070504	Premature Discharge of Battery	21/07/2021	21/07/2021
									A070909	Oversensing	21/07/2021	21/07/2021
					CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	21/07/2021	21/07/2021
									A070909	Oversensing	21/07/2021	21/07/2021

		F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	21/07/2021	21/07/2021
				A070909	Oversensing	21/07/2021	21/07/2021
		F1903	Device Explantation	A070504	Premature Discharge of Battery	21/07/2021	21/07/2021
				A070909	Oversensing	21/07/2021	21/07/2021
CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	21/07/2021	21/07/2021
				A070909	Oversensing	21/07/2021	21/07/2021
		F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	21/07/2021	21/07/2021
				A070909	Oversensing	21/07/2021	21/07/2021
		F1903	Device Explantation	A070504	Premature Discharge of Battery	21/07/2021	21/07/2021
				A070909	Oversensing	21/07/2021	21/07/2021
CT233	Surgical	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	21/07/2021	21/07/2021
				A070909	Oversensing	21/07/2021	21/07/2021
		F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	21/07/2021	21/07/2021
				A070909	Oversensing	21/07/2021	21/07/2021
		F1903	Device Explantation	A070504	Premature Discharge of Battery	21/07/2021	21/07/2021
				A070909	Oversensing	21/07/2021	21/07/2021
CT2406	Implantable	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	21/07/2021	21/07/2021
				A070909	Oversensing	21/07/2021	21/07/2021
		F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	21/07/2021	21/07/2021
				A070909	Oversensing	21/07/2021	21/07/2021
		F1903	Device Explantation	A070504	Premature Discharge of Battery	21/07/2021	21/07/2021
				A070909	Oversensing	21/07/2021	21/07/2021
CT2407	Active implantable	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	21/07/2021	21/07/2021
				A070909	Oversensing	21/07/2021	21/07/2021
		F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	21/07/2021	21/07/2021
				A070909	Oversensing	21/07/2021	21/07/2021
		F1903	Device Explantation	A070504	Premature Discharge of Battery	21/07/2021	21/07/2021
				A070909	Oversensing	21/07/2021	21/07/2021
CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	21/07/2021	21/07/2021
				A070909	Oversensing	21/07/2021	21/07/2021
		F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	21/07/2021	21/07/2021
				A070909	Oversensing	21/07/2021	21/07/2021
		F1903	Device Explantation	A070504	Premature Discharge of Battery	21/07/2021	21/07/2021
				A070909	Oversensing	21/07/2021	21/07/2021
CT2727	Active implant devices/ancillaries	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	21/07/2021	21/07/2021
				A070909	Oversensing	21/07/2021	21/07/2021
		F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	21/07/2021	21/07/2021
				A070909	Oversensing	21/07/2021	21/07/2021
		F1903	Device Explantation	A070504	Premature Discharge of Battery	21/07/2021	21/07/2021
				A070909	Oversensing	21/07/2021	21/07/2021
CT2750	Implantable cardiac electrical stimulators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	21/07/2021	21/07/2021
				A070909	Oversensing	21/07/2021	21/07/2021
		F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	21/07/2021	21/07/2021
				A070909	Oversensing	21/07/2021	21/07/2021
		F1903	Device Explantation	A070504	Premature Discharge of Battery	21/07/2021	21/07/2021
				A070909	Oversensing	21/07/2021	21/07/2021
CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	21/07/2021	21/07/2021
				A070909	Oversensing	21/07/2021	21/07/2021
		F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	21/07/2021	21/07/2021
				A070909	Oversensing	21/07/2021	21/07/2021
		F1903	Device Explantation	A070504	Premature Discharge of Battery	21/07/2021	21/07/2021
				A070909	Oversensing	21/07/2021	21/07/2021
CT2985	Active	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	21/07/2021	21/07/2021
				A070909	Oversensing	21/07/2021	21/07/2021
		F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	21/07/2021	21/07/2021
				A070909	Oversensing	21/07/2021	21/07/2021
		F1903	Device Explantation	A070504	Premature Discharge of Battery	21/07/2021	21/07/2021
				A070909	Oversensing	21/07/2021	21/07/2021
CT3001	Implantable metal	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	21/07/2021	21/07/2021
				A070909	Oversensing	21/07/2021	21/07/2021
		F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	21/07/2021	21/07/2021
				A070909	Oversensing	21/07/2021	21/07/2021
		F1903	Device Explantation	A070504	Premature Discharge of Battery	21/07/2021	21/07/2021
				A070909	Oversensing	21/07/2021	21/07/2021
CT3002	Implantable synthetic polymer	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	21/07/2021	21/07/2021
				A070909	Oversensing	21/07/2021	21/07/2021
		F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	21/07/2021	21/07/2021
				A070909	Oversensing	21/07/2021	21/07/2021
		F1903	Device Explantation	A070504	Premature Discharge of Battery	21/07/2021	21/07/2021
				A070909	Oversensing	21/07/2021	21/07/2021
CT3006	Implantable inorganic material	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	21/07/2021	21/07/2021
				A070909	Oversensing	21/07/2021	21/07/2021
		F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	21/07/2021	21/07/2021
				A070909	Oversensing	21/07/2021	21/07/2021
		F1903	Device Explantation	A070504	Premature Discharge of Battery	21/07/2021	21/07/2021
				A070909	Oversensing	21/07/2021	21/07/2021
CT314	Electrically powered	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	21/07/2021	21/07/2021
				A070909	Oversensing	21/07/2021	21/07/2021
		F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	21/07/2021	21/07/2021
				A070909	Oversensing	21/07/2021	21/07/2021
		F1903	Device Explantation	A070504	Premature Discharge of Battery	21/07/2021	21/07/2021
				A070909	Oversensing	21/07/2021	21/07/2021
CT321	Long-term surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	21/07/2021	21/07/2021
				A070909	Oversensing	21/07/2021	21/07/2021
		F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	21/07/2021	21/07/2021
				A070909	Oversensing	21/07/2021	21/07/2021
		F1903	Device Explantation	A070504	Premature Discharge of Battery	21/07/2021	21/07/2021
				A070909	Oversensing	21/07/2021	21/07/2021

					CT449	Generators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	21/07/2021	21/07/2021
									A070909	Oversensing	21/07/2021	21/07/2021
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	21/07/2021	21/07/2021
									A070909	Oversensing	21/07/2021	21/07/2021
							F1903	Device Explantation	A070504	Premature Discharge of Battery	21/07/2021	21/07/2021
									A070909	Oversensing	21/07/2021	21/07/2021
					CT981	Single use	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	21/07/2021	21/07/2021
									A070909	Oversensing	21/07/2021	21/07/2021
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	21/07/2021	21/07/2021
									A070909	Oversensing	21/07/2021	21/07/2021
							F1903	Device Explantation	A070504	Premature Discharge of Battery	21/07/2021	21/07/2021
									A070909	Oversensing	21/07/2021	21/07/2021
					CT983	Surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	21/07/2021	21/07/2021
									A070909	Oversensing	21/07/2021	21/07/2021
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	21/07/2021	21/07/2021
									A070909	Oversensing	21/07/2021	21/07/2021
							F1903	Device Explantation	A070504	Premature Discharge of Battery	21/07/2021	21/07/2021
									A070909	Oversensing	21/07/2021	21/07/2021
6248	2021	All other reportable incidents	47265	Dual-chamber implantable pacemaker, rate-responsive	CT141	Electrodes/Leads and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem	04/08/2021	04/08/2021
									A070504	Premature Discharge of Battery	04/08/2021	04/08/2021
							F08	Hospitalization or Prolonged Hospitalization	A0705	Battery Problem	04/08/2021	04/08/2021
									A070504	Premature Discharge of Battery	04/08/2021	04/08/2021
							F1905	Device Revision or Replacement	A0705	Battery Problem	04/08/2021	04/08/2021
									A070504	Premature Discharge of Battery	04/08/2021	04/08/2021
					CT145	Cardiovascular devices	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem	04/08/2021	04/08/2021
									A070504	Premature Discharge of Battery	04/08/2021	04/08/2021
							F08	Hospitalization or Prolonged Hospitalization	A0705	Battery Problem	04/08/2021	04/08/2021
									A070504	Premature Discharge of Battery	04/08/2021	04/08/2021
							F1905	Device Revision or Replacement	A0705	Battery Problem	04/08/2021	04/08/2021
									A070504	Premature Discharge of Battery	04/08/2021	04/08/2021
					CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem	04/08/2021	04/08/2021
									A070504	Premature Discharge of Battery	04/08/2021	04/08/2021
							F08	Hospitalization or Prolonged Hospitalization	A0705	Battery Problem	04/08/2021	04/08/2021
									A070504	Premature Discharge of Battery	04/08/2021	04/08/2021
							F1905	Device Revision or Replacement	A0705	Battery Problem	04/08/2021	04/08/2021
									A070504	Premature Discharge of Battery	04/08/2021	04/08/2021
					CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem	04/08/2021	04/08/2021
									A070504	Premature Discharge of Battery	04/08/2021	04/08/2021
							F08	Hospitalization or Prolonged Hospitalization	A0705	Battery Problem	04/08/2021	04/08/2021
									A070504	Premature Discharge of Battery	04/08/2021	04/08/2021
							F1905	Device Revision or Replacement	A0705	Battery Problem	04/08/2021	04/08/2021
									A070504	Premature Discharge of Battery	04/08/2021	04/08/2021
					CT233	Surgical	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem	04/08/2021	04/08/2021
									A070504	Premature Discharge of Battery	04/08/2021	04/08/2021
							F08	Hospitalization or Prolonged Hospitalization	A0705	Battery Problem	04/08/2021	04/08/2021
									A070504	Premature Discharge of Battery	04/08/2021	04/08/2021
							F1905	Device Revision or Replacement	A0705	Battery Problem	04/08/2021	04/08/2021
									A070504	Premature Discharge of Battery	04/08/2021	04/08/2021
					CT2406	Implantable	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem	04/08/2021	04/08/2021
									A070504	Premature Discharge of Battery	04/08/2021	04/08/2021
							F08	Hospitalization or Prolonged Hospitalization	A0705	Battery Problem	04/08/2021	04/08/2021
									A070504	Premature Discharge of Battery	04/08/2021	04/08/2021
							F1905	Device Revision or Replacement	A0705	Battery Problem	04/08/2021	04/08/2021
									A070504	Premature Discharge of Battery	04/08/2021	04/08/2021
					CT2407	Active implantable	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem	04/08/2021	04/08/2021
									A070504	Premature Discharge of Battery	04/08/2021	04/08/2021
							F08	Hospitalization or Prolonged Hospitalization	A0705	Battery Problem	04/08/2021	04/08/2021
									A070504	Premature Discharge of Battery	04/08/2021	04/08/2021
							F1905	Device Revision or Replacement	A0705	Battery Problem	04/08/2021	04/08/2021
									A070504	Premature Discharge of Battery	04/08/2021	04/08/2021
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem	04/08/2021	04/08/2021
									A070504	Premature Discharge of Battery	04/08/2021	04/08/2021
							F08	Hospitalization or Prolonged Hospitalization	A0705	Battery Problem	04/08/2021	04/08/2021
									A070504	Premature Discharge of Battery	04/08/2021	04/08/2021
							F1905	Device Revision or Replacement	A0705	Battery Problem	04/08/2021	04/08/2021
									A070504	Premature Discharge of Battery	04/08/2021	04/08/2021
					CT2727	Active implant devices/ancillaries	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem	04/08/2021	04/08/2021
									A070504	Premature Discharge of Battery	04/08/2021	04/08/2021
							F08	Hospitalization or Prolonged Hospitalization	A0705	Battery Problem	04/08/2021	04/08/2021
									A070504	Premature Discharge of Battery	04/08/2021	04/08/2021
							F1905	Device Revision or Replacement	A0705	Battery Problem	04/08/2021	04/08/2021
									A070504	Premature Discharge of Battery	04/08/2021	04/08/2021
					CT2750	Implantable cardiac electrical stimulators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem	04/08/2021	04/08/2021
									A070504	Premature Discharge of Battery	04/08/2021	04/08/2021
							F08	Hospitalization or Prolonged Hospitalization	A0705	Battery Problem	04/08/2021	04/08/2021
									A070504	Premature Discharge of Battery	04/08/2021	04/08/2021
							F1905	Device Revision or Replacement	A0705	Battery Problem	04/08/2021	04/08/2021
									A070504	Premature Discharge of Battery	04/08/2021	04/08/2021
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem	04/08/2021	04/08/2021
									A070504	Premature Discharge of Battery	04/08/2021	04/08/2021
							F08	Hospitalization or Prolonged Hospitalization	A0705	Battery Problem	04/08/2021	04/08/2021
									A070504	Premature Discharge of Battery	04/08/2021	04/08/2021
							F1905	Device Revision or Replacement	A0705	Battery Problem	04/08/2021	04/08/2021
									A070504	Premature Discharge of Battery	04/08/2021	04/08/2021
					CT2985	Active	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem	04/08/2021	04/08/2021
									A070504	Premature Discharge of Battery	04/08/2021	04/08/2021
							F08	Hospitalization or Prolonged Hospitalization	A0705	Battery Problem	04/08/2021	04/08/2021
									A070504	Premature Discharge of Battery	04/08/2021	04/08/2021

							F1905	Device Revision or Replacement	A0705	Battery Problem	04/08/2021	04/08/2021
									A070504	Premature Discharge of Battery	04/08/2021	04/08/2021
					CT3001	Implantable metal	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem	04/08/2021	04/08/2021
									A070504	Premature Discharge of Battery	04/08/2021	04/08/2021
							F08	Hospitalization or Prolonged Hospitalization	A0705	Battery Problem	04/08/2021	04/08/2021
									A070504	Premature Discharge of Battery	04/08/2021	04/08/2021
							F1905	Device Revision or Replacement	A0705	Battery Problem	04/08/2021	04/08/2021
									A070504	Premature Discharge of Battery	04/08/2021	04/08/2021
					CT3002	Implantable synthetic polymer	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem	04/08/2021	04/08/2021
									A070504	Premature Discharge of Battery	04/08/2021	04/08/2021
							F08	Hospitalization or Prolonged Hospitalization	A0705	Battery Problem	04/08/2021	04/08/2021
									A070504	Premature Discharge of Battery	04/08/2021	04/08/2021
							F1905	Device Revision or Replacement	A0705	Battery Problem	04/08/2021	04/08/2021
									A070504	Premature Discharge of Battery	04/08/2021	04/08/2021
					CT3006	Implantable inorganic material	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem	04/08/2021	04/08/2021
									A070504	Premature Discharge of Battery	04/08/2021	04/08/2021
							F08	Hospitalization or Prolonged Hospitalization	A0705	Battery Problem	04/08/2021	04/08/2021
									A070504	Premature Discharge of Battery	04/08/2021	04/08/2021
							F1905	Device Revision or Replacement	A0705	Battery Problem	04/08/2021	04/08/2021
									A070504	Premature Discharge of Battery	04/08/2021	04/08/2021
					CT314	Electrically powered	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem	04/08/2021	04/08/2021
									A070504	Premature Discharge of Battery	04/08/2021	04/08/2021
							F08	Hospitalization or Prolonged Hospitalization	A0705	Battery Problem	04/08/2021	04/08/2021
									A070504	Premature Discharge of Battery	04/08/2021	04/08/2021
							F1905	Device Revision or Replacement	A0705	Battery Problem	04/08/2021	04/08/2021
									A070504	Premature Discharge of Battery	04/08/2021	04/08/2021
					CT321	Long-term surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem	04/08/2021	04/08/2021
									A070504	Premature Discharge of Battery	04/08/2021	04/08/2021
							F08	Hospitalization or Prolonged Hospitalization	A0705	Battery Problem	04/08/2021	04/08/2021
									A070504	Premature Discharge of Battery	04/08/2021	04/08/2021
							F1905	Device Revision or Replacement	A0705	Battery Problem	04/08/2021	04/08/2021
									A070504	Premature Discharge of Battery	04/08/2021	04/08/2021
					CT449	Generators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem	04/08/2021	04/08/2021
									A070504	Premature Discharge of Battery	04/08/2021	04/08/2021
							F08	Hospitalization or Prolonged Hospitalization	A0705	Battery Problem	04/08/2021	04/08/2021
									A070504	Premature Discharge of Battery	04/08/2021	04/08/2021
							F1905	Device Revision or Replacement	A0705	Battery Problem	04/08/2021	04/08/2021
									A070504	Premature Discharge of Battery	04/08/2021	04/08/2021
					CT981	Single use	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem	04/08/2021	04/08/2021
									A070504	Premature Discharge of Battery	04/08/2021	04/08/2021
							F08	Hospitalization or Prolonged Hospitalization	A0705	Battery Problem	04/08/2021	04/08/2021
									A070504	Premature Discharge of Battery	04/08/2021	04/08/2021
							F1905	Device Revision or Replacement	A0705	Battery Problem	04/08/2021	04/08/2021
									A070504	Premature Discharge of Battery	04/08/2021	04/08/2021
					CT983	Surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem	04/08/2021	04/08/2021
									A070504	Premature Discharge of Battery	04/08/2021	04/08/2021
							F08	Hospitalization or Prolonged Hospitalization	A0705	Battery Problem	04/08/2021	04/08/2021
									A070504	Premature Discharge of Battery	04/08/2021	04/08/2021
							F1905	Device Revision or Replacement	A0705	Battery Problem	04/08/2021	04/08/2021
									A070504	Premature Discharge of Battery	04/08/2021	04/08/2021
6249	2022	All other reportable incidents	47265	Dual-chamber implantable pacemaker, rate-responsive	CT141	Electrodes/Leads and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	11/03/2022	11/03/2022
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	11/03/2022	11/03/2022
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	11/03/2022	11/03/2022
					CT145	Cardiovascular devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	11/03/2022	11/03/2022
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	11/03/2022	11/03/2022
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	11/03/2022	11/03/2022
					CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	11/03/2022	11/03/2022
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	11/03/2022	11/03/2022
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	11/03/2022	11/03/2022
					CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	11/03/2022	11/03/2022
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	11/03/2022	11/03/2022
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	11/03/2022	11/03/2022
					CT233	Surgical	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	11/03/2022	11/03/2022
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	11/03/2022	11/03/2022
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	11/03/2022	11/03/2022
					CT2406	Implantable	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	11/03/2022	11/03/2022
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	11/03/2022	11/03/2022
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	11/03/2022	11/03/2022
					CT2407	Active implantable	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	11/03/2022	11/03/2022
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	11/03/2022	11/03/2022
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	11/03/2022	11/03/2022
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	11/03/2022	11/03/2022
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	11/03/2022	11/03/2022
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	11/03/2022	11/03/2022
					CT2727	Active implant devices/ancillaries	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	11/03/2022	11/03/2022
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	11/03/2022	11/03/2022
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	11/03/2022	11/03/2022
					CT2750	Implantable cardiac electrical stimulators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	11/03/2022	11/03/2022
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	11/03/2022	11/03/2022
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	11/03/2022	11/03/2022
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	11/03/2022	11/03/2022
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	11/03/2022	11/03/2022
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	11/03/2022	11/03/2022
					CT2985	Active	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	11/03/2022	11/03/2022
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	11/03/2022	11/03/2022
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	11/03/2022	11/03/2022
					CT3001	Implantable metal	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	11/03/2022	11/03/2022
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	11/03/2022	11/03/2022

						F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	11/03/2022	11/03/2022
				CT3002	Implantable synthetic polymer	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	11/03/2022	11/03/2022
						F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	11/03/2022	11/03/2022
						F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	11/03/2022	11/03/2022
				CT3006	Implantable inorganic material	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	11/03/2022	11/03/2022
						F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	11/03/2022	11/03/2022
						F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	11/03/2022	11/03/2022
				CT314	Electrically powered	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	11/03/2022	11/03/2022
						F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	11/03/2022	11/03/2022
						F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	11/03/2022	11/03/2022
				CT321	Long-term surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	11/03/2022	11/03/2022
						F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	11/03/2022	11/03/2022
						F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	11/03/2022	11/03/2022
				CT449	Generators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	11/03/2022	11/03/2022
						F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	11/03/2022	11/03/2022
						F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	11/03/2022	11/03/2022
				CT981	Single use	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	11/03/2022	11/03/2022
						F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	11/03/2022	11/03/2022
						F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	11/03/2022	11/03/2022
				CT983	Surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	11/03/2022	11/03/2022
						F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	11/03/2022	11/03/2022
						F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	11/03/2022	11/03/2022
6250	2022	All other reportable incidents	47267	Single-chamber implantable pacemaker, rate-responsive	CT141	Electrodes/Leads and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	16/12/2021	16/12/2021
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	16/12/2021	16/12/2021
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	16/12/2021	16/12/2021
					CT145	Cardiovascular devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	16/12/2021	16/12/2021
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	16/12/2021	16/12/2021
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	16/12/2021	16/12/2021
					CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	16/12/2021	16/12/2021
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	16/12/2021	16/12/2021
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	16/12/2021	16/12/2021
					CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	16/12/2021	16/12/2021
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	16/12/2021	16/12/2021
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	16/12/2021	16/12/2021
					CT233	Surgical	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	16/12/2021	16/12/2021
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	16/12/2021	16/12/2021
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	16/12/2021	16/12/2021
					CT2406	Implantable	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	16/12/2021	16/12/2021
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	16/12/2021	16/12/2021
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	16/12/2021	16/12/2021
					CT2407	Active implantable	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	16/12/2021	16/12/2021
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	16/12/2021	16/12/2021
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	16/12/2021	16/12/2021
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	16/12/2021	16/12/2021
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	16/12/2021	16/12/2021
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	16/12/2021	16/12/2021
					CT2727	Active implant devices/ancillaries	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	16/12/2021	16/12/2021
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	16/12/2021	16/12/2021
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	16/12/2021	16/12/2021
					CT2750	Implantable cardiac electrical stimulators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	16/12/2021	16/12/2021
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	16/12/2021	16/12/2021
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	16/12/2021	16/12/2021
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	16/12/2021	16/12/2021
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	16/12/2021	16/12/2021
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	16/12/2021	16/12/2021
					CT2985	Active	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	16/12/2021	16/12/2021
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	16/12/2021	16/12/2021
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	16/12/2021	16/12/2021
					CT3001	Implantable metal	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	16/12/2021	16/12/2021
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	16/12/2021	16/12/2021
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	16/12/2021	16/12/2021
					CT3002	Implantable synthetic polymer	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	16/12/2021	16/12/2021
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	16/12/2021	16/12/2021
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	16/12/2021	16/12/2021
					CT3006	Implantable inorganic material	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	16/12/2021	16/12/2021
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	16/12/2021	16/12/2021
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	16/12/2021	16/12/2021
					CT314	Electrically powered	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	16/12/2021	16/12/2021
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	16/12/2021	16/12/2021
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	16/12/2021	16/12/2021
					CT321	Long-term surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	16/12/2021	16/12/2021
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	16/12/2021	16/12/2021
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	16/12/2021	16/12/2021
					CT449	Generators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	16/12/2021	16/12/2021
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	16/12/2021	16/12/2021
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	16/12/2021	16/12/2021
					CT981	Single use	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	16/12/2021	16/12/2021
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	16/12/2021	16/12/2021
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	16/12/2021	16/12/2021
					CT983	Surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	16/12/2021	16/12/2021
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	16/12/2021	16/12/2021
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	16/12/2021	16/12/2021
6251	2022	All other reportable incidents	47263	Cardiac resynchronization therapy implantable pacemaker	CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A0908	Incorrect, Inadequate or Imprecise Result or Readings	03/03/2022	03/03/2022
							F08	Hospitalization or Prolonged Hospitalization	A0908	Incorrect, Inadequate or Imprecise Result or Readings	03/03/2022	03/03/2022
							F1905	Device Revision or Replacement	A0908	Incorrect, Inadequate or Imprecise Result or Readings	03/03/2022	03/03/2022

					CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A0908	Incorrect, Inadequate or Imprecise Result or Readings	03/03/2022	03/03/2022
							F08	Hospitalization or Prolonged Hospitalization	A0908	Incorrect, Inadequate or Imprecise Result or Readings	03/03/2022	03/03/2022
							F1905	Device Revision or Replacement	A0908	Incorrect, Inadequate or Imprecise Result or Readings	03/03/2022	03/03/2022
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A0908	Incorrect, Inadequate or Imprecise Result or Readings	03/03/2022	03/03/2022
							F08	Hospitalization or Prolonged Hospitalization	A0908	Incorrect, Inadequate or Imprecise Result or Readings	03/03/2022	03/03/2022
							F1905	Device Revision or Replacement	A0908	Incorrect, Inadequate or Imprecise Result or Readings	03/03/2022	03/03/2022
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A0908	Incorrect, Inadequate or Imprecise Result or Readings	03/03/2022	03/03/2022
							F08	Hospitalization or Prolonged Hospitalization	A0908	Incorrect, Inadequate or Imprecise Result or Readings	03/03/2022	03/03/2022
							F1905	Device Revision or Replacement	A0908	Incorrect, Inadequate or Imprecise Result or Readings	03/03/2022	03/03/2022
6252	2021	All other reportable incidents	47263	Cardiac resynchronization therapy implantable pacemaker	CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
6253	2021	Death	35223	Endocardial/interventricular septal pacing lead	CT1684	Cardiac pacemakers and associated devices	E2402	Appropriate Clinical Signs, Symptoms and Conditions Term/Code not Available	A0401	Break		
									A070909	Oversensing		
									A090801	Signal Artifact/Noise		
							F02	Death	A0401	Break		
									A070909	Oversensing		
									A090801	Signal Artifact/Noise		
6254	2021	All other reportable incidents	17882	Physiologic-monitoring defibrillation system	CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A0801	Failure to Calibrate		
							F27	Problem Identified During Non-Clinical Procedure	A0801	Failure to Calibrate		
6255	2021	Death	47270	Cardiac resynchronization therapy implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	10/08/2021	10/08/2021
									A071201	Failure to Convert Rhythm	10/08/2021	10/08/2021
									A072202	Low Impedance	10/08/2021	10/08/2021
									A0908	Incorrect, Inadequate or Imprecise Result or Readings	10/08/2021	10/08/2021
							F02	Death	A070504	Premature Discharge of Battery	10/08/2021	10/08/2021
									A071201	Failure to Convert Rhythm	10/08/2021	10/08/2021
									A072202	Low Impedance	10/08/2021	10/08/2021
									A0908	Incorrect, Inadequate or Imprecise Result or Readings	10/08/2021	10/08/2021
6256	2021	Death	36241	Pacemaker/defibrillator lead	CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A071201	Failure to Convert Rhythm		
									A072202	Low Impedance		
							F02	Death	A071201	Failure to Convert Rhythm		
									A072202	Low Impedance		
6257	2022	Unanticipated serious deterioration in state of health	47265	Dual-chamber implantable pacemaker, rate-responsive	CT141	Electrodes/Leads and associated devices	E2401	Insufficient Information	A24	No Apparent Device Problem		
							F02	Death	A24	No Apparent Device Problem		
					CT145	Cardiovascular devices	E2401	Insufficient Information	A24	No Apparent Device Problem		
							F02	Death	A24	No Apparent Device Problem		
					CT1684	Cardiac pacemakers and associated devices	E2401	Insufficient Information	A24	No Apparent Device Problem		
							F02	Death	A24	No Apparent Device Problem		
					CT1685	Cardiac pacemakers	E2401	Insufficient Information	A24	No Apparent Device Problem		
							F02	Death	A24	No Apparent Device Problem		
					CT233	Surgical	E2401	Insufficient Information	A24	No Apparent Device Problem		
							F02	Death	A24	No Apparent Device Problem		
					CT2406	Implantable	E2401	Insufficient Information	A24	No Apparent Device Problem		
							F02	Death	A24	No Apparent Device Problem		
					CT2407	Active implantable	E2401	Insufficient Information	A24	No Apparent Device Problem		
							F02	Death	A24	No Apparent Device Problem		
					CT257	Implantable cardiac pacemakers	E2401	Insufficient Information	A24	No Apparent Device Problem		
							F02	Death	A24	No Apparent Device Problem		
					CT2727	Active implant devices/ancillaries	E2401	Insufficient Information	A24	No Apparent Device Problem		
							F02	Death	A24	No Apparent Device Problem		
					CT2750	Implantable cardiac electrical stimulators and associated devices	E2401	Insufficient Information	A24	No Apparent Device Problem		
							F02	Death	A24	No Apparent Device Problem		
					CT2866	Implantable cardiac electrical stimulators	E2401	Insufficient Information	A24	No Apparent Device Problem		
							F02	Death	A24	No Apparent Device Problem		
					CT2985	Active	E2401	Insufficient Information	A24	No Apparent Device Problem		
							F02	Death	A24	No Apparent Device Problem		
					CT3001	Implantable metal	E2401	Insufficient Information	A24	No Apparent Device Problem		
							F02	Death	A24	No Apparent Device Problem		
					CT3002	Implantable synthetic polymer	E2401	Insufficient Information	A24	No Apparent Device Problem		
							F02	Death	A24	No Apparent Device Problem		
					CT3006	Implantable inorganic material	E2401	Insufficient Information	A24	No Apparent Device Problem		
							F02	Death	A24	No Apparent Device Problem		
					CT314	Electrically powered	E2401	Insufficient Information	A24	No Apparent Device Problem		
							F02	Death	A24	No Apparent Device Problem		
					CT321	Long-term surgical invasive	E2401	Insufficient Information	A24	No Apparent Device Problem		
							F02	Death	A24	No Apparent Device Problem		
					CT449	Generators and associated devices	E2401	Insufficient Information	A24	No Apparent Device Problem		
							F02	Death	A24	No Apparent Device Problem		
					CT981	Single use	E2401	Insufficient Information	A24	No Apparent Device Problem		
							F02	Death	A24	No Apparent Device Problem		
					CT983	Surgical invasive	E2401	Insufficient Information	A24	No Apparent Device Problem		
							F02	Death	A24	No Apparent Device Problem		
6258	2022	All other reportable incidents	17882	Physiologic-monitoring defibrillation system	CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A1205	Loose or Intermittent Connection		
							F27	Problem Identified During Non-Clinical Procedure	A1205	Loose or Intermittent Connection		

6259	2021	All other reportable incidents	35852	Single-chamber implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E0601	Arrhythmia	A070603	Failure to Charge	12/11/2021	12/11/2021
									A071201	Failure to Convert Rhythm	12/11/2021	12/11/2021
									A071301	Failure to Deliver Shock/Stimulation	12/11/2021	12/11/2021
									A090805	Incorrect Measurement	12/11/2021	12/11/2021
							E2403	No Clinical Signs, Symptoms or Conditions	A070603	Failure to Charge	12/11/2021	12/11/2021
									A071201	Failure to Convert Rhythm	12/11/2021	12/11/2021
									A071301	Failure to Deliver Shock/Stimulation	12/11/2021	12/11/2021
									A090805	Incorrect Measurement	12/11/2021	12/11/2021
							F1901	Additional Surgery	A070603	Failure to Charge	12/11/2021	12/11/2021
									A071201	Failure to Convert Rhythm	12/11/2021	12/11/2021
									A071301	Failure to Deliver Shock/Stimulation	12/11/2021	12/11/2021
									A090805	Incorrect Measurement	12/11/2021	12/11/2021
6260	2022	All other reportable incidents	35852	Single-chamber implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A090801	Signal Artifact/Noise		
							F26	No Health Consequences or Impact	A090801	Signal Artifact/Noise		
6261	2022	All other reportable incidents	47265	Dual-chamber implantable pacemaker, rate-responsive	CT141	Electrodes/Leads and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT145	Cardiovascular devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT233	Surgical	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT2406	Implantable	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT2407	Active implantable	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT2727	Active implant devices/ancillaries	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT2750	Implantable cardiac electrical stimulators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT2985	Active	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT3001	Implantable metal	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT3002	Implantable synthetic polymer	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT3006	Implantable inorganic material	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT314	Electrically powered	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT321	Long-term surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT449	Generators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT981	Single use	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT983	Surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
6262	2022	All other reportable incidents	47265	Dual-chamber implantable pacemaker, rate-responsive	CT141	Electrodes/Leads and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	19/03/2022	19/03/2022
							F1901	Additional Surgery	A070504	Premature Discharge of Battery	19/03/2022	19/03/2022
					CT145	Cardiovascular devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	19/03/2022	19/03/2022
							F1901	Additional Surgery	A070504	Premature Discharge of Battery	19/03/2022	19/03/2022
					CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	19/03/2022	19/03/2022
							F1901	Additional Surgery	A070504	Premature Discharge of Battery	19/03/2022	19/03/2022
					CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	19/03/2022	19/03/2022
							F1901	Additional Surgery	A070504	Premature Discharge of Battery	19/03/2022	19/03/2022
					CT233	Surgical	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	19/03/2022	19/03/2022
							F1901	Additional Surgery	A070504	Premature Discharge of Battery	19/03/2022	19/03/2022
					CT2406	Implantable	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	19/03/2022	19/03/2022
							F1901	Additional Surgery	A070504	Premature Discharge of Battery	19/03/2022	19/03/2022
					CT2407	Active implantable	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	19/03/2022	19/03/2022
							F1901	Additional Surgery	A070504	Premature Discharge of Battery	19/03/2022	19/03/2022
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	19/03/2022	19/03/2022
							F1901	Additional Surgery	A070504	Premature Discharge of Battery	19/03/2022	19/03/2022
					CT2727	Active implant devices/ancillaries	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	19/03/2022	19/03/2022
							F1901	Additional Surgery	A070504	Premature Discharge of Battery	19/03/2022	19/03/2022
					CT2750	Implantable cardiac electrical stimulators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	19/03/2022	19/03/2022
							F1901	Additional Surgery	A070504	Premature Discharge of Battery	19/03/2022	19/03/2022
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	19/03/2022	19/03/2022
							F1901	Additional Surgery	A070504	Premature Discharge of Battery	19/03/2022	19/03/2022
					CT2985	Active	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	19/03/2022	19/03/2022
							F1901	Additional Surgery	A070504	Premature Discharge of Battery	19/03/2022	19/03/2022
					CT3001	Implantable metal	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	19/03/2022	19/03/2022
							F1901	Additional Surgery	A070504	Premature Discharge of Battery	19/03/2022	19/03/2022
					CT3002	Implantable synthetic polymer	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	19/03/2022	19/03/2022
							F1901	Additional Surgery	A070504	Premature Discharge of Battery	19/03/2022	19/03/2022
					CT3006	Implantable inorganic material	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	19/03/2022	19/03/2022
							F1901	Additional Surgery	A070504	Premature Discharge of Battery	19/03/2022	19/03/2022
					CT314	Electrically powered	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	19/03/2022	19/03/2022
							F1901	Additional Surgery	A070504	Premature Discharge of Battery	19/03/2022	19/03/2022
					CT321	Long-term surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	19/03/2022	19/03/2022
							F1901	Additional Surgery	A070504	Premature Discharge of Battery	19/03/2022	19/03/2022

					CT449	Generators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	19/03/2022	19/03/2022
							F1901	Additional Surgery	A070504	Premature Discharge of Battery	19/03/2022	19/03/2022
					CT981	Single use	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	19/03/2022	19/03/2022
							F1901	Additional Surgery	A070504	Premature Discharge of Battery	19/03/2022	19/03/2022
					CT983	Surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	19/03/2022	19/03/2022
							F1901	Additional Surgery	A070504	Premature Discharge of Battery	19/03/2022	19/03/2022
6263	2021	All other reportable incidents	47265	Dual-chamber implantable pacemaker, rate-responsive	CT141	Electrodes/Leads and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	05/08/2021	05/08/2021
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery	05/08/2021	05/08/2021
					CT145	Cardiovascular devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	05/08/2021	05/08/2021
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery	05/08/2021	05/08/2021
					CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	05/08/2021	05/08/2021
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery	05/08/2021	05/08/2021
					CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	05/08/2021	05/08/2021
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery	05/08/2021	05/08/2021
					CT233	Surgical	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	05/08/2021	05/08/2021
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery	05/08/2021	05/08/2021
					CT2406	Implantable	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	05/08/2021	05/08/2021
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery	05/08/2021	05/08/2021
					CT2407	Active implantable	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	05/08/2021	05/08/2021
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery	05/08/2021	05/08/2021
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	05/08/2021	05/08/2021
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery	05/08/2021	05/08/2021
					CT2727	Active implant devices/ancillaries	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	05/08/2021	05/08/2021
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery	05/08/2021	05/08/2021
					CT2750	Implantable cardiac electrical stimulators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	05/08/2021	05/08/2021
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery	05/08/2021	05/08/2021
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	05/08/2021	05/08/2021
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery	05/08/2021	05/08/2021
					CT2985	Active	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	05/08/2021	05/08/2021
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery	05/08/2021	05/08/2021
					CT3001	Implantable metal	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	05/08/2021	05/08/2021
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery	05/08/2021	05/08/2021
					CT3002	Implantable synthetic polymer	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	05/08/2021	05/08/2021
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery	05/08/2021	05/08/2021
					CT3006	Implantable inorganic material	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	05/08/2021	05/08/2021
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery	05/08/2021	05/08/2021
					CT314	Electrically powered	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	05/08/2021	05/08/2021
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery	05/08/2021	05/08/2021
					CT321	Long-term surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	05/08/2021	05/08/2021
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery	05/08/2021	05/08/2021
					CT449	Generators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	05/08/2021	05/08/2021
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery	05/08/2021	05/08/2021
					CT981	Single use	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	05/08/2021	05/08/2021
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery	05/08/2021	05/08/2021
					CT983	Surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	05/08/2021	05/08/2021
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery	05/08/2021	05/08/2021
6264	2021	All other reportable incidents	36241	Pacemaker/defibrillator lead	CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A051201	Device Dislodged or Dislocated		
									A070101	Failure to Capture		
							F26	No Health Consequences or Impact	A051201	Device Dislodged or Dislocated		
									A070101	Failure to Capture		
6265	2021	All other reportable incidents	35852	Single-chamber implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem	10/02/2022	10/02/2022
									A070504	Premature Discharge of Battery	10/02/2022	10/02/2022
							F08	Hospitalization or Prolonged Hospitalization	A0705	Battery Problem	10/02/2022	10/02/2022
									A070504	Premature Discharge of Battery	10/02/2022	10/02/2022
							F1905	Device Revision or Replacement	A0705	Battery Problem	10/02/2022	10/02/2022
									A070504	Premature Discharge of Battery	10/02/2022	10/02/2022
6266	2021	All other reportable incidents	35852	Single-chamber implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem	29/09/2021	29/09/2021
									A070504	Premature Discharge of Battery	29/09/2021	29/09/2021
							F08	Hospitalization or Prolonged Hospitalization	A0705	Battery Problem	29/09/2021	29/09/2021
									A070504	Premature Discharge of Battery	29/09/2021	29/09/2021
							F1905	Device Revision or Replacement	A0705	Battery Problem	29/09/2021	29/09/2021
									A070504	Premature Discharge of Battery	29/09/2021	29/09/2021
6267	2021	All other reportable incidents	36241	Pacemaker/defibrillator lead	CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A051201	Device Dislodged or Dislocated	27/07/2021	27/07/2021
									A070102	High Capture Threshold	27/07/2021	27/07/2021
							F08	Hospitalization or Prolonged Hospitalization	A051201	Device Dislodged or Dislocated	27/07/2021	27/07/2021
									A070102	High Capture Threshold	27/07/2021	27/07/2021
							F1905	Device Revision or Replacement	A051201	Device Dislodged or Dislocated	27/07/2021	27/07/2021
									A070102	High Capture Threshold	27/07/2021	27/07/2021
6268	2021	All other reportable incidents	47263	Cardiac resynchronization therapy implantable pacemaker	CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	03/08/2021	03/08/2021
							F08	Hospitalization or Prolonged Hospitalization	A24	No Apparent Device Problem	03/08/2021	03/08/2021
							F1905	Device Revision or Replacement	A24	No Apparent Device Problem	03/08/2021	03/08/2021
					CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	03/08/2021	03/08/2021
							F08	Hospitalization or Prolonged Hospitalization	A24	No Apparent Device Problem	03/08/2021	03/08/2021
							F1905	Device Revision or Replacement	A24	No Apparent Device Problem	03/08/2021	03/08/2021
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	03/08/2021	03/08/2021
							F08	Hospitalization or Prolonged Hospitalization	A24	No Apparent Device Problem	03/08/2021	03/08/2021
							F1905	Device Revision or Replacement	A24	No Apparent Device Problem	03/08/2021	03/08/2021
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	03/08/2021	03/08/2021
							F08	Hospitalization or Prolonged Hospitalization	A24	No Apparent Device Problem	03/08/2021	03/08/2021
							F1905	Device Revision or Replacement	A24	No Apparent Device Problem	03/08/2021	03/08/2021
6269	2021	All other reportable incidents	47265	Dual-chamber implantable pacemaker, rate-responsive	CT141	Electrodes/Leads and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	27/07/2021	27/07/2021
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	27/07/2021	27/07/2021
							F1903	Device Explantation	A070504	Premature Discharge of Battery	27/07/2021	27/07/2021
					CT145	Cardiovascular devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	27/07/2021	27/07/2021
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	27/07/2021	27/07/2021
							F1903	Device Explantation	A070504	Premature Discharge of Battery	27/07/2021	27/07/2021
					CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	27/07/2021	27/07/2021
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	27/07/2021	27/07/2021

							F1903	Device Explantation	A070504	Premature Discharge of Battery	27/07/2021	27/07/2021
					CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	27/07/2021	27/07/2021
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	27/07/2021	27/07/2021
							F1903	Device Explantation	A070504	Premature Discharge of Battery	27/07/2021	27/07/2021
					CT233	Surgical	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	27/07/2021	27/07/2021
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	27/07/2021	27/07/2021
							F1903	Device Explantation	A070504	Premature Discharge of Battery	27/07/2021	27/07/2021
					CT2406	Implantable	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	27/07/2021	27/07/2021
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	27/07/2021	27/07/2021
							F1903	Device Explantation	A070504	Premature Discharge of Battery	27/07/2021	27/07/2021
					CT2407	Active implantable	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	27/07/2021	27/07/2021
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	27/07/2021	27/07/2021
							F1903	Device Explantation	A070504	Premature Discharge of Battery	27/07/2021	27/07/2021
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	27/07/2021	27/07/2021
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	27/07/2021	27/07/2021
							F1903	Device Explantation	A070504	Premature Discharge of Battery	27/07/2021	27/07/2021
					CT2727	Active implant devices/ancillaries	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	27/07/2021	27/07/2021
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	27/07/2021	27/07/2021
							F1903	Device Explantation	A070504	Premature Discharge of Battery	27/07/2021	27/07/2021
					CT2750	Implantable cardiac electrical stimulators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	27/07/2021	27/07/2021
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	27/07/2021	27/07/2021
							F1903	Device Explantation	A070504	Premature Discharge of Battery	27/07/2021	27/07/2021
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	27/07/2021	27/07/2021
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	27/07/2021	27/07/2021
							F1903	Device Explantation	A070504	Premature Discharge of Battery	27/07/2021	27/07/2021
					CT2985	Active	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	27/07/2021	27/07/2021
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	27/07/2021	27/07/2021
							F1903	Device Explantation	A070504	Premature Discharge of Battery	27/07/2021	27/07/2021
					CT3001	Implantable metal	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	27/07/2021	27/07/2021
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	27/07/2021	27/07/2021
							F1903	Device Explantation	A070504	Premature Discharge of Battery	27/07/2021	27/07/2021
					CT3002	Implantable synthetic polymer	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	27/07/2021	27/07/2021
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	27/07/2021	27/07/2021
							F1903	Device Explantation	A070504	Premature Discharge of Battery	27/07/2021	27/07/2021
					CT3006	Implantable inorganic material	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	27/07/2021	27/07/2021
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	27/07/2021	27/07/2021
							F1903	Device Explantation	A070504	Premature Discharge of Battery	27/07/2021	27/07/2021
					CT314	Electrically powered	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	27/07/2021	27/07/2021
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	27/07/2021	27/07/2021
							F1903	Device Explantation	A070504	Premature Discharge of Battery	27/07/2021	27/07/2021
					CT321	Long-term surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	27/07/2021	27/07/2021
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	27/07/2021	27/07/2021
							F1903	Device Explantation	A070504	Premature Discharge of Battery	27/07/2021	27/07/2021
					CT449	Generators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	27/07/2021	27/07/2021
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	27/07/2021	27/07/2021
							F1903	Device Explantation	A070504	Premature Discharge of Battery	27/07/2021	27/07/2021
					CT981	Single use	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	27/07/2021	27/07/2021
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	27/07/2021	27/07/2021
							F1903	Device Explantation	A070504	Premature Discharge of Battery	27/07/2021	27/07/2021
					CT983	Surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	27/07/2021	27/07/2021
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	27/07/2021	27/07/2021
							F1903	Device Explantation	A070504	Premature Discharge of Battery	27/07/2021	27/07/2021
6270	2021	All other reportable incidents	35852	Single-chamber implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem	14/09/2021	14/09/2021
									A070504	Premature Discharge of Battery	14/09/2021	14/09/2021
							F08	Hospitalization or Prolonged Hospitalization	A0705	Battery Problem	14/09/2021	14/09/2021
									A070504	Premature Discharge of Battery	14/09/2021	14/09/2021
							F1905	Device Revision or Replacement	A0705	Battery Problem	14/09/2021	14/09/2021
									A070504	Premature Discharge of Battery	14/09/2021	14/09/2021
6271	2021	All other reportable incidents	47267	Single-chamber implantable pacemaker, rate-responsive	CT141	Electrodes/Leads and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	17/08/2021	17/08/2021
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	17/08/2021	17/08/2021
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	17/08/2021	17/08/2021
					CT145	Cardiovascular devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	17/08/2021	17/08/2021
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	17/08/2021	17/08/2021
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	17/08/2021	17/08/2021
					CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	17/08/2021	17/08/2021
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	17/08/2021	17/08/2021
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	17/08/2021	17/08/2021
					CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	17/08/2021	17/08/2021
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	17/08/2021	17/08/2021
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	17/08/2021	17/08/2021
					CT233	Surgical	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	17/08/2021	17/08/2021
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	17/08/2021	17/08/2021
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	17/08/2021	17/08/2021
					CT2406	Implantable	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	17/08/2021	17/08/2021
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	17/08/2021	17/08/2021
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	17/08/2021	17/08/2021
					CT2407	Active implantable	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	17/08/2021	17/08/2021
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	17/08/2021	17/08/2021
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	17/08/2021	17/08/2021
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	17/08/2021	17/08/2021
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	17/08/2021	17/08/2021
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	17/08/2021	17/08/2021
					CT2727	Active implant devices/ancillaries	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	17/08/2021	17/08/2021
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	17/08/2021	17/08/2021
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	17/08/2021	17/08/2021
					CT2750	Implantable cardiac electrical stimulators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	17/08/2021	17/08/2021
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	17/08/2021	17/08/2021
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	17/08/2021	17/08/2021

					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	17/08/2021	17/08/2021
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	17/08/2021	17/08/2021
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	17/08/2021	17/08/2021
					CT2985	Active	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	17/08/2021	17/08/2021
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	17/08/2021	17/08/2021
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	17/08/2021	17/08/2021
					CT3001	Implantable metal	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	17/08/2021	17/08/2021
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	17/08/2021	17/08/2021
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	17/08/2021	17/08/2021
					CT3002	Implantable synthetic polymer	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	17/08/2021	17/08/2021
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	17/08/2021	17/08/2021
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	17/08/2021	17/08/2021
					CT3006	Implantable inorganic material	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	17/08/2021	17/08/2021
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	17/08/2021	17/08/2021
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	17/08/2021	17/08/2021
					CT314	Electrically powered	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	17/08/2021	17/08/2021
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	17/08/2021	17/08/2021
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	17/08/2021	17/08/2021
					CT321	Long-term surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	17/08/2021	17/08/2021
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	17/08/2021	17/08/2021
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	17/08/2021	17/08/2021
					CT449	Generators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	17/08/2021	17/08/2021
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	17/08/2021	17/08/2021
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	17/08/2021	17/08/2021
					CT981	Single use	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	17/08/2021	17/08/2021
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	17/08/2021	17/08/2021
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	17/08/2021	17/08/2021
					CT983	Surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	17/08/2021	17/08/2021
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	17/08/2021	17/08/2021
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	17/08/2021	17/08/2021
6272	2021		17882	Physiologic-monitoring defibrillation system	CT1684	Cardiac pacemakers and associated devices						
6273	2021	All other reportable incidents	47263	Cardiac resynchronization therapy implantable pacemaker	CT1684	Cardiac pacemakers and associated devices	E012207	Twitching	A0908	Incorrect, Inadequate or Imprecise Result or Readings	03/08/2021	03/08/2021
							E2403	No Clinical Signs, Symptoms or Conditions	A0908	Incorrect, Inadequate or Imprecise Result or Readings	03/08/2021	03/08/2021
							F08	Hospitalization or Prolonged Hospitalization	A0908	Incorrect, Inadequate or Imprecise Result or Readings	03/08/2021	03/08/2021
							F1905	Device Revision or Replacement	A0908	Incorrect, Inadequate or Imprecise Result or Readings	03/08/2021	03/08/2021
					CT1685	Cardiac pacemakers	E012207	Twitching	A0908	Incorrect, Inadequate or Imprecise Result or Readings	03/08/2021	03/08/2021
							E2403	No Clinical Signs, Symptoms or Conditions	A0908	Incorrect, Inadequate or Imprecise Result or Readings	03/08/2021	03/08/2021
							F08	Hospitalization or Prolonged Hospitalization	A0908	Incorrect, Inadequate or Imprecise Result or Readings	03/08/2021	03/08/2021
							F1905	Device Revision or Replacement	A0908	Incorrect, Inadequate or Imprecise Result or Readings	03/08/2021	03/08/2021
					CT257	Implantable cardiac pacemakers	E012207	Twitching	A0908	Incorrect, Inadequate or Imprecise Result or Readings	03/08/2021	03/08/2021
							E2403	No Clinical Signs, Symptoms or Conditions	A0908	Incorrect, Inadequate or Imprecise Result or Readings	03/08/2021	03/08/2021
							F08	Hospitalization or Prolonged Hospitalization	A0908	Incorrect, Inadequate or Imprecise Result or Readings	03/08/2021	03/08/2021
							F1905	Device Revision or Replacement	A0908	Incorrect, Inadequate or Imprecise Result or Readings	03/08/2021	03/08/2021
					CT2866	Implantable cardiac electrical stimulators	E012207	Twitching	A0908	Incorrect, Inadequate or Imprecise Result or Readings	03/08/2021	03/08/2021
							E2403	No Clinical Signs, Symptoms or Conditions	A0908	Incorrect, Inadequate or Imprecise Result or Readings	03/08/2021	03/08/2021
							F08	Hospitalization or Prolonged Hospitalization	A0908	Incorrect, Inadequate or Imprecise Result or Readings	03/08/2021	03/08/2021
							F1905	Device Revision or Replacement	A0908	Incorrect, Inadequate or Imprecise Result or Readings	03/08/2021	03/08/2021
6274	2021	All other reportable incidents	35852	Single-chamber implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem	05/08/2021	05/08/2021
									A070504	Premature Discharge of Battery	05/08/2021	05/08/2021
							F08	Hospitalization or Prolonged Hospitalization	A0705	Battery Problem	05/08/2021	05/08/2021
									A070504	Premature Discharge of Battery	05/08/2021	05/08/2021
							F1905	Device Revision or Replacement	A0705	Battery Problem	05/08/2021	05/08/2021
									A070504	Premature Discharge of Battery	05/08/2021	05/08/2021
6275	2021	All other reportable incidents	47265	Dual-chamber implantable pacemaker, rate-responsive	CT141	Electrodes/Leads and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	23/07/2021	23/07/2021
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	23/07/2021	23/07/2021
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	23/07/2021	23/07/2021
					CT145	Cardiovascular devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	23/07/2021	23/07/2021
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	23/07/2021	23/07/2021
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	23/07/2021	23/07/2021
					CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	23/07/2021	23/07/2021
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	23/07/2021	23/07/2021
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	23/07/2021	23/07/2021
					CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	23/07/2021	23/07/2021
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	23/07/2021	23/07/2021
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	23/07/2021	23/07/2021
					CT233	Surgical	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	23/07/2021	23/07/2021
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	23/07/2021	23/07/2021
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	23/07/2021	23/07/2021
					CT2406	Implantable	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	23/07/2021	23/07/2021
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	23/07/2021	23/07/2021
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	23/07/2021	23/07/2021
					CT2407	Active implantable	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	23/07/2021	23/07/2021
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	23/07/2021	23/07/2021
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	23/07/2021	23/07/2021
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	23/07/2021	23/07/2021
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	23/07/2021	23/07/2021

							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	23/07/2021	23/07/2021
					CT2727	Active implant devices/ancillaries	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	23/07/2021	23/07/2021
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	23/07/2021	23/07/2021
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	23/07/2021	23/07/2021
					CT2750	Implantable cardiac electrical stimulators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	23/07/2021	23/07/2021
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	23/07/2021	23/07/2021
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	23/07/2021	23/07/2021
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	23/07/2021	23/07/2021
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	23/07/2021	23/07/2021
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	23/07/2021	23/07/2021
					CT2985	Active	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	23/07/2021	23/07/2021
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	23/07/2021	23/07/2021
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	23/07/2021	23/07/2021
					CT3001	Implantable metal	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	23/07/2021	23/07/2021
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	23/07/2021	23/07/2021
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	23/07/2021	23/07/2021
					CT3002	Implantable synthetic polymer	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	23/07/2021	23/07/2021
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	23/07/2021	23/07/2021
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	23/07/2021	23/07/2021
					CT3006	Implantable inorganic material	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	23/07/2021	23/07/2021
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	23/07/2021	23/07/2021
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	23/07/2021	23/07/2021
					CT314	Electrically powered	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	23/07/2021	23/07/2021
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	23/07/2021	23/07/2021
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	23/07/2021	23/07/2021
					CT321	Long-term surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	23/07/2021	23/07/2021
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	23/07/2021	23/07/2021
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	23/07/2021	23/07/2021
					CT449	Generators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	23/07/2021	23/07/2021
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	23/07/2021	23/07/2021
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	23/07/2021	23/07/2021
					CT981	Single use	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	23/07/2021	23/07/2021
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	23/07/2021	23/07/2021
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	23/07/2021	23/07/2021
					CT983	Surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	23/07/2021	23/07/2021
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	23/07/2021	23/07/2021
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	23/07/2021	23/07/2021
8276	2021	All other reportable incidents	47265	Dual-chamber implantable pacemaker, rate-responsive	CT141	Electrodes/Leads and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem	29/07/2021	29/07/2021
									A070504	Premature Discharge of Battery	29/07/2021	29/07/2021
							F08	Hospitalization or Prolonged Hospitalization	A0705	Battery Problem	29/07/2021	29/07/2021
									A070504	Premature Discharge of Battery	29/07/2021	29/07/2021
							F1905	Device Revision or Replacement	A0705	Battery Problem	29/07/2021	29/07/2021
									A070504	Premature Discharge of Battery	29/07/2021	29/07/2021
					CT145	Cardiovascular devices	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem	29/07/2021	29/07/2021
									A070504	Premature Discharge of Battery	29/07/2021	29/07/2021
							F08	Hospitalization or Prolonged Hospitalization	A0705	Battery Problem	29/07/2021	29/07/2021
									A070504	Premature Discharge of Battery	29/07/2021	29/07/2021
							F1905	Device Revision or Replacement	A0705	Battery Problem	29/07/2021	29/07/2021
									A070504	Premature Discharge of Battery	29/07/2021	29/07/2021
					CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem	29/07/2021	29/07/2021
									A070504	Premature Discharge of Battery	29/07/2021	29/07/2021
							F08	Hospitalization or Prolonged Hospitalization	A0705	Battery Problem	29/07/2021	29/07/2021
									A070504	Premature Discharge of Battery	29/07/2021	29/07/2021
							F1905	Device Revision or Replacement	A0705	Battery Problem	29/07/2021	29/07/2021
									A070504	Premature Discharge of Battery	29/07/2021	29/07/2021
					CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem	29/07/2021	29/07/2021
									A070504	Premature Discharge of Battery	29/07/2021	29/07/2021
							F08	Hospitalization or Prolonged Hospitalization	A0705	Battery Problem	29/07/2021	29/07/2021
									A070504	Premature Discharge of Battery	29/07/2021	29/07/2021
							F1905	Device Revision or Replacement	A0705	Battery Problem	29/07/2021	29/07/2021
									A070504	Premature Discharge of Battery	29/07/2021	29/07/2021
					CT233	Surgical	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem	29/07/2021	29/07/2021
									A070504	Premature Discharge of Battery	29/07/2021	29/07/2021
							F08	Hospitalization or Prolonged Hospitalization	A0705	Battery Problem	29/07/2021	29/07/2021
									A070504	Premature Discharge of Battery	29/07/2021	29/07/2021
							F1905	Device Revision or Replacement	A0705	Battery Problem	29/07/2021	29/07/2021
									A070504	Premature Discharge of Battery	29/07/2021	29/07/2021
					CT2406	Implantable	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem	29/07/2021	29/07/2021
									A070504	Premature Discharge of Battery	29/07/2021	29/07/2021
							F08	Hospitalization or Prolonged Hospitalization	A0705	Battery Problem	29/07/2021	29/07/2021
									A070504	Premature Discharge of Battery	29/07/2021	29/07/2021
							F1905	Device Revision or Replacement	A0705	Battery Problem	29/07/2021	29/07/2021
									A070504	Premature Discharge of Battery	29/07/2021	29/07/2021
					CT2407	Active implantable	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem	29/07/2021	29/07/2021
									A070504	Premature Discharge of Battery	29/07/2021	29/07/2021
							F08	Hospitalization or Prolonged Hospitalization	A0705	Battery Problem	29/07/2021	29/07/2021
									A070504	Premature Discharge of Battery	29/07/2021	29/07/2021
							F1905	Device Revision or Replacement	A0705	Battery Problem	29/07/2021	29/07/2021
									A070504	Premature Discharge of Battery	29/07/2021	29/07/2021
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem	29/07/2021	29/07/2021
									A070504	Premature Discharge of Battery	29/07/2021	29/07/2021
							F08	Hospitalization or Prolonged Hospitalization	A0705	Battery Problem	29/07/2021	29/07/2021
									A070504	Premature Discharge of Battery	29/07/2021	29/07/2021
							F1905	Device Revision or Replacement	A0705	Battery Problem	29/07/2021	29/07/2021
									A070504	Premature Discharge of Battery	29/07/2021	29/07/2021
					CT2727	Active implant devices/ancillaries	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem	29/07/2021	29/07/2021
									A070504	Premature Discharge of Battery	29/07/2021	29/07/2021
							F08	Hospitalization or Prolonged Hospitalization	A0705	Battery Problem	29/07/2021	29/07/2021

									A070504	Premature Discharge of Battery	29/07/2021	29/07/2021
							F1905	Device Revision or Replacement	A0705	Battery Problem	29/07/2021	29/07/2021
									A070504	Premature Discharge of Battery	29/07/2021	29/07/2021
					CT2750	Implantable cardiac electrical stimulators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem	29/07/2021	29/07/2021
									A070504	Premature Discharge of Battery	29/07/2021	29/07/2021
							F08	Hospitalization or Prolonged Hospitalization	A0705	Battery Problem	29/07/2021	29/07/2021
									A070504	Premature Discharge of Battery	29/07/2021	29/07/2021
							F1905	Device Revision or Replacement	A0705	Battery Problem	29/07/2021	29/07/2021
									A070504	Premature Discharge of Battery	29/07/2021	29/07/2021
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem	29/07/2021	29/07/2021
									A070504	Premature Discharge of Battery	29/07/2021	29/07/2021
							F08	Hospitalization or Prolonged Hospitalization	A0705	Battery Problem	29/07/2021	29/07/2021
									A070504	Premature Discharge of Battery	29/07/2021	29/07/2021
							F1905	Device Revision or Replacement	A0705	Battery Problem	29/07/2021	29/07/2021
									A070504	Premature Discharge of Battery	29/07/2021	29/07/2021
					CT2985	Active	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem	29/07/2021	29/07/2021
									A070504	Premature Discharge of Battery	29/07/2021	29/07/2021
							F08	Hospitalization or Prolonged Hospitalization	A0705	Battery Problem	29/07/2021	29/07/2021
									A070504	Premature Discharge of Battery	29/07/2021	29/07/2021
							F1905	Device Revision or Replacement	A0705	Battery Problem	29/07/2021	29/07/2021
									A070504	Premature Discharge of Battery	29/07/2021	29/07/2021
					CT3001	Implantable metal	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem	29/07/2021	29/07/2021
									A070504	Premature Discharge of Battery	29/07/2021	29/07/2021
							F08	Hospitalization or Prolonged Hospitalization	A0705	Battery Problem	29/07/2021	29/07/2021
									A070504	Premature Discharge of Battery	29/07/2021	29/07/2021
							F1905	Device Revision or Replacement	A0705	Battery Problem	29/07/2021	29/07/2021
									A070504	Premature Discharge of Battery	29/07/2021	29/07/2021
					CT3002	Implantable synthetic polymer	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem	29/07/2021	29/07/2021
									A070504	Premature Discharge of Battery	29/07/2021	29/07/2021
							F08	Hospitalization or Prolonged Hospitalization	A0705	Battery Problem	29/07/2021	29/07/2021
									A070504	Premature Discharge of Battery	29/07/2021	29/07/2021
							F1905	Device Revision or Replacement	A0705	Battery Problem	29/07/2021	29/07/2021
									A070504	Premature Discharge of Battery	29/07/2021	29/07/2021
					CT3006	Implantable inorganic material	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem	29/07/2021	29/07/2021
									A070504	Premature Discharge of Battery	29/07/2021	29/07/2021
							F08	Hospitalization or Prolonged Hospitalization	A0705	Battery Problem	29/07/2021	29/07/2021
									A070504	Premature Discharge of Battery	29/07/2021	29/07/2021
							F1905	Device Revision or Replacement	A0705	Battery Problem	29/07/2021	29/07/2021
									A070504	Premature Discharge of Battery	29/07/2021	29/07/2021
					CT314	Electrically powered	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem	29/07/2021	29/07/2021
									A070504	Premature Discharge of Battery	29/07/2021	29/07/2021
							F08	Hospitalization or Prolonged Hospitalization	A0705	Battery Problem	29/07/2021	29/07/2021
									A070504	Premature Discharge of Battery	29/07/2021	29/07/2021
							F1905	Device Revision or Replacement	A0705	Battery Problem	29/07/2021	29/07/2021
									A070504	Premature Discharge of Battery	29/07/2021	29/07/2021
					CT321	Long-term surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem	29/07/2021	29/07/2021
									A070504	Premature Discharge of Battery	29/07/2021	29/07/2021
							F08	Hospitalization or Prolonged Hospitalization	A0705	Battery Problem	29/07/2021	29/07/2021
									A070504	Premature Discharge of Battery	29/07/2021	29/07/2021
							F1905	Device Revision or Replacement	A0705	Battery Problem	29/07/2021	29/07/2021
									A070504	Premature Discharge of Battery	29/07/2021	29/07/2021
					CT449	Generators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem	29/07/2021	29/07/2021
									A070504	Premature Discharge of Battery	29/07/2021	29/07/2021
							F08	Hospitalization or Prolonged Hospitalization	A0705	Battery Problem	29/07/2021	29/07/2021
									A070504	Premature Discharge of Battery	29/07/2021	29/07/2021
							F1905	Device Revision or Replacement	A0705	Battery Problem	29/07/2021	29/07/2021
									A070504	Premature Discharge of Battery	29/07/2021	29/07/2021
					CT981	Single use	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem	29/07/2021	29/07/2021
									A070504	Premature Discharge of Battery	29/07/2021	29/07/2021
							F08	Hospitalization or Prolonged Hospitalization	A0705	Battery Problem	29/07/2021	29/07/2021
									A070504	Premature Discharge of Battery	29/07/2021	29/07/2021
							F1905	Device Revision or Replacement	A0705	Battery Problem	29/07/2021	29/07/2021
									A070504	Premature Discharge of Battery	29/07/2021	29/07/2021
					CT983	Surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem	29/07/2021	29/07/2021
									A070504	Premature Discharge of Battery	29/07/2021	29/07/2021
							F08	Hospitalization or Prolonged Hospitalization	A0705	Battery Problem	29/07/2021	29/07/2021
									A070504	Premature Discharge of Battery	29/07/2021	29/07/2021
							F1905	Device Revision or Replacement	A0705	Battery Problem	29/07/2021	29/07/2021
									A070504	Premature Discharge of Battery	29/07/2021	29/07/2021
6277	2021	All other reportable incidents	35852	Single-chamber implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E2104	Electric Shock	A070909	Oversensing		
									A071302	Inappropriate/Inadequate Shock/Stimulation		
									A1301	Failure to Read Input Signal		
							E2403	No Clinical Signs, Symptoms or Conditions	A070909	Oversensing		
									A071302	Inappropriate/Inadequate Shock/Stimulation		
									A1301	Failure to Read Input Signal		
							F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A070909	Oversensing		
									A071302	Inappropriate/Inadequate Shock/Stimulation		
									A1301	Failure to Read Input Signal		
6278	2021	All other reportable incidents	35852	Single-chamber implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E2104	Electric Shock	A070909	Oversensing		
									A071302	Inappropriate/Inadequate Shock/Stimulation		
							E2403	No Clinical Signs, Symptoms or Conditions	A070909	Oversensing		
									A071302	Inappropriate/Inadequate Shock/Stimulation		
							F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A070909	Oversensing		
									A071302	Inappropriate/Inadequate Shock/Stimulation		
6279	2022	All other reportable incidents	63188	Left-ventricular wireless pacing system sensor/transmitter	CT141	Electrodes/Leads and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A1101	Application Network Problem		
					CT145	Cardiovascular devices	E2403	No Clinical Signs, Symptoms or Conditions	A1101	Application Network Problem		
					CT1577	Wireless monitoring system transmitters	E2403	No Clinical Signs, Symptoms or Conditions	A1101	Application Network Problem		
					CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A1101	Application Network Problem		

					CT2317	Transmitters/Receivers/Transceivers	E2403	No Clinical Signs, Symptoms or Conditions	A1101	Application Network Problem		
					CT233	Surgical	E2403	No Clinical Signs, Symptoms or Conditions	A1101	Application Network Problem		
					CT2406	Implantable	E2403	No Clinical Signs, Symptoms or Conditions	A1101	Application Network Problem		
					CT2407	Active implantable	E2403	No Clinical Signs, Symptoms or Conditions	A1101	Application Network Problem		
					CT255	Transducers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A1101	Application Network Problem		
					CT2726	Physiological monitoring/measuring devices	E2403	No Clinical Signs, Symptoms or Conditions	A1101	Application Network Problem		
					CT2727	Active implant devices/ancillaries	E2403	No Clinical Signs, Symptoms or Conditions	A1101	Application Network Problem		
					CT2750	Implantable cardiac electrical stimulators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A1101	Application Network Problem		
					CT2885	Physiological monitoring/measuring system transmitters/receivers/transceivers	E2403	No Clinical Signs, Symptoms or Conditions	A1101	Application Network Problem		
					CT2985	Active	E2403	No Clinical Signs, Symptoms or Conditions	A1101	Application Network Problem		
					CT3001	Implantable metal	E2403	No Clinical Signs, Symptoms or Conditions	A1101	Application Network Problem		
					CT3002	Implantable synthetic polymer	E2403	No Clinical Signs, Symptoms or Conditions	A1101	Application Network Problem		
					CT3006	Implantable inorganic material	E2403	No Clinical Signs, Symptoms or Conditions	A1101	Application Network Problem		
					CT314	Electrically powered	E2403	No Clinical Signs, Symptoms or Conditions	A1101	Application Network Problem		
					CT321	Long-term surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A1101	Application Network Problem		
					CT384	Sensors and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A1101	Application Network Problem		
					CT449	Generators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A1101	Application Network Problem		
					CT981	Single use	E2403	No Clinical Signs, Symptoms or Conditions	A1101	Application Network Problem		
					CT983	Surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A1101	Application Network Problem		
					CT994	Sensors	E2403	No Clinical Signs, Symptoms or Conditions	A1101	Application Network Problem		
6280	2022	All other reportable incidents	63188	Left-ventricular wireless pacing system sensor/transmitter	CT141	Electrodes/Leads and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A1101	Application Network Problem		
					CT145	Cardiovascular devices	E2403	No Clinical Signs, Symptoms or Conditions	A1101	Application Network Problem		
					CT1577	Wireless monitoring system transmitters	E2403	No Clinical Signs, Symptoms or Conditions	A1101	Application Network Problem		
					CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A1101	Application Network Problem		
					CT2317	Transmitters/Receivers/Transceivers	E2403	No Clinical Signs, Symptoms or Conditions	A1101	Application Network Problem		
					CT233	Surgical	E2403	No Clinical Signs, Symptoms or Conditions	A1101	Application Network Problem		
					CT2406	Implantable	E2403	No Clinical Signs, Symptoms or Conditions	A1101	Application Network Problem		
					CT2407	Active implantable	E2403	No Clinical Signs, Symptoms or Conditions	A1101	Application Network Problem		
					CT255	Transducers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A1101	Application Network Problem		
					CT2726	Physiological monitoring/measuring devices	E2403	No Clinical Signs, Symptoms or Conditions	A1101	Application Network Problem		
					CT2727	Active implant devices/ancillaries	E2403	No Clinical Signs, Symptoms or Conditions	A1101	Application Network Problem		
					CT2750	Implantable cardiac electrical stimulators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A1101	Application Network Problem		
					CT2885	Physiological monitoring/measuring system transmitters/receivers/transceivers	E2403	No Clinical Signs, Symptoms or Conditions	A1101	Application Network Problem		
					CT2985	Active	E2403	No Clinical Signs, Symptoms or Conditions	A1101	Application Network Problem		
					CT3001	Implantable metal	E2403	No Clinical Signs, Symptoms or Conditions	A1101	Application Network Problem		
					CT3002	Implantable synthetic polymer	E2403	No Clinical Signs, Symptoms or Conditions	A1101	Application Network Problem		
					CT3006	Implantable inorganic material	E2403	No Clinical Signs, Symptoms or Conditions	A1101	Application Network Problem		
					CT314	Electrically powered	E2403	No Clinical Signs, Symptoms or Conditions	A1101	Application Network Problem		
					CT321	Long-term surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A1101	Application Network Problem		
					CT384	Sensors and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A1101	Application Network Problem		
					CT449	Generators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A1101	Application Network Problem		
					CT981	Single use	E2403	No Clinical Signs, Symptoms or Conditions	A1101	Application Network Problem		
					CT983	Surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A1101	Application Network Problem		
					CT994	Sensors	E2403	No Clinical Signs, Symptoms or Conditions	A1101	Application Network Problem		
6281	2021	All other reportable incidents	47265	Dual-chamber implantable pacemaker, rate-responsive	CT141	Electrodes/Leads and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT145	Cardiovascular devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT233	Surgical	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT2406	Implantable	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT2407	Active implantable	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT2727	Active implant devices/ancillaries	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT2750	Implantable cardiac electrical stimulators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT2985	Active	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT3001	Implantable metal	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT3002	Implantable synthetic polymer	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT3006	Implantable inorganic material	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT314	Electrically powered	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT321	Long-term surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT449	Generators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT981	Single use	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT983	Surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
6282	2021	All other reportable incidents	47265	Dual-chamber implantable pacemaker, rate-	CT141	Electrodes/Leads and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem	27/10/2021	27/10/2021

				responsive					A070504	Premature Discharge of Battery	27/10/2021	27/10/2021
							F08	Hospitalization or Prolonged Hospitalization	A0705	Battery Problem	27/10/2021	27/10/2021
									A070504	Premature Discharge of Battery	27/10/2021	27/10/2021
							F1905	Device Revision or Replacement	A0705	Battery Problem	27/10/2021	27/10/2021
									A070504	Premature Discharge of Battery	27/10/2021	27/10/2021
					CT145	Cardiovascular devices	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem	27/10/2021	27/10/2021
									A070504	Premature Discharge of Battery	27/10/2021	27/10/2021
							F08	Hospitalization or Prolonged Hospitalization	A0705	Battery Problem	27/10/2021	27/10/2021
									A070504	Premature Discharge of Battery	27/10/2021	27/10/2021
							F1905	Device Revision or Replacement	A0705	Battery Problem	27/10/2021	27/10/2021
									A070504	Premature Discharge of Battery	27/10/2021	27/10/2021
					CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem	27/10/2021	27/10/2021
									A070504	Premature Discharge of Battery	27/10/2021	27/10/2021
							F08	Hospitalization or Prolonged Hospitalization	A0705	Battery Problem	27/10/2021	27/10/2021
									A070504	Premature Discharge of Battery	27/10/2021	27/10/2021
							F1905	Device Revision or Replacement	A0705	Battery Problem	27/10/2021	27/10/2021
									A070504	Premature Discharge of Battery	27/10/2021	27/10/2021
					CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem	27/10/2021	27/10/2021
									A070504	Premature Discharge of Battery	27/10/2021	27/10/2021
							F08	Hospitalization or Prolonged Hospitalization	A0705	Battery Problem	27/10/2021	27/10/2021
									A070504	Premature Discharge of Battery	27/10/2021	27/10/2021
							F1905	Device Revision or Replacement	A0705	Battery Problem	27/10/2021	27/10/2021
									A070504	Premature Discharge of Battery	27/10/2021	27/10/2021
					CT233	Surgical	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem	27/10/2021	27/10/2021
									A070504	Premature Discharge of Battery	27/10/2021	27/10/2021
							F08	Hospitalization or Prolonged Hospitalization	A0705	Battery Problem	27/10/2021	27/10/2021
									A070504	Premature Discharge of Battery	27/10/2021	27/10/2021
							F1905	Device Revision or Replacement	A0705	Battery Problem	27/10/2021	27/10/2021
									A070504	Premature Discharge of Battery	27/10/2021	27/10/2021
					CT2406	Implantable	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem	27/10/2021	27/10/2021
									A070504	Premature Discharge of Battery	27/10/2021	27/10/2021
							F08	Hospitalization or Prolonged Hospitalization	A0705	Battery Problem	27/10/2021	27/10/2021
									A070504	Premature Discharge of Battery	27/10/2021	27/10/2021
							F1905	Device Revision or Replacement	A0705	Battery Problem	27/10/2021	27/10/2021
									A070504	Premature Discharge of Battery	27/10/2021	27/10/2021
					CT2407	Active implantable	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem	27/10/2021	27/10/2021
									A070504	Premature Discharge of Battery	27/10/2021	27/10/2021
							F08	Hospitalization or Prolonged Hospitalization	A0705	Battery Problem	27/10/2021	27/10/2021
									A070504	Premature Discharge of Battery	27/10/2021	27/10/2021
							F1905	Device Revision or Replacement	A0705	Battery Problem	27/10/2021	27/10/2021
									A070504	Premature Discharge of Battery	27/10/2021	27/10/2021
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem	27/10/2021	27/10/2021
									A070504	Premature Discharge of Battery	27/10/2021	27/10/2021
							F08	Hospitalization or Prolonged Hospitalization	A0705	Battery Problem	27/10/2021	27/10/2021
									A070504	Premature Discharge of Battery	27/10/2021	27/10/2021
							F1905	Device Revision or Replacement	A0705	Battery Problem	27/10/2021	27/10/2021
									A070504	Premature Discharge of Battery	27/10/2021	27/10/2021
					CT2727	Active implant devices/ancillaries	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem	27/10/2021	27/10/2021
									A070504	Premature Discharge of Battery	27/10/2021	27/10/2021
							F08	Hospitalization or Prolonged Hospitalization	A0705	Battery Problem	27/10/2021	27/10/2021
									A070504	Premature Discharge of Battery	27/10/2021	27/10/2021
							F1905	Device Revision or Replacement	A0705	Battery Problem	27/10/2021	27/10/2021
									A070504	Premature Discharge of Battery	27/10/2021	27/10/2021
					CT2750	Implantable cardiac electrical stimulators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem	27/10/2021	27/10/2021
									A070504	Premature Discharge of Battery	27/10/2021	27/10/2021
							F08	Hospitalization or Prolonged Hospitalization	A0705	Battery Problem	27/10/2021	27/10/2021
									A070504	Premature Discharge of Battery	27/10/2021	27/10/2021
							F1905	Device Revision or Replacement	A0705	Battery Problem	27/10/2021	27/10/2021
									A070504	Premature Discharge of Battery	27/10/2021	27/10/2021
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem	27/10/2021	27/10/2021
									A070504	Premature Discharge of Battery	27/10/2021	27/10/2021
							F08	Hospitalization or Prolonged Hospitalization	A0705	Battery Problem	27/10/2021	27/10/2021
									A070504	Premature Discharge of Battery	27/10/2021	27/10/2021
							F1905	Device Revision or Replacement	A0705	Battery Problem	27/10/2021	27/10/2021
									A070504	Premature Discharge of Battery	27/10/2021	27/10/2021
					CT2985	Active	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem	27/10/2021	27/10/2021
									A070504	Premature Discharge of Battery	27/10/2021	27/10/2021
							F08	Hospitalization or Prolonged Hospitalization	A0705	Battery Problem	27/10/2021	27/10/2021
									A070504	Premature Discharge of Battery	27/10/2021	27/10/2021
							F1905	Device Revision or Replacement	A0705	Battery Problem	27/10/2021	27/10/2021
									A070504	Premature Discharge of Battery	27/10/2021	27/10/2021
					CT3001	Implantable metal	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem	27/10/2021	27/10/2021
									A070504	Premature Discharge of Battery	27/10/2021	27/10/2021
							F08	Hospitalization or Prolonged Hospitalization	A0705	Battery Problem	27/10/2021	27/10/2021
									A070504	Premature Discharge of Battery	27/10/2021	27/10/2021
							F1905	Device Revision or Replacement	A0705	Battery Problem	27/10/2021	27/10/2021
									A070504	Premature Discharge of Battery	27/10/2021	27/10/2021
					CT3002	Implantable synthetic polymer	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem	27/10/2021	27/10/2021
									A070504	Premature Discharge of Battery	27/10/2021	27/10/2021
							F08	Hospitalization or Prolonged Hospitalization	A0705	Battery Problem	27/10/2021	27/10/2021
									A070504	Premature Discharge of Battery	27/10/2021	27/10/2021
							F1905	Device Revision or Replacement	A0705	Battery Problem	27/10/2021	27/10/2021
									A070504	Premature Discharge of Battery	27/10/2021	27/10/2021
					CT3006	Implantable inorganic material	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem	27/10/2021	27/10/2021
									A070504	Premature Discharge of Battery	27/10/2021	27/10/2021
							F08	Hospitalization or Prolonged Hospitalization	A0705	Battery Problem	27/10/2021	27/10/2021
									A070504	Premature Discharge of Battery	27/10/2021	27/10/2021
							F1905	Device Revision or Replacement	A0705	Battery Problem	27/10/2021	27/10/2021

									A070504	Premature Discharge of Battery	27/10/2021	27/10/2021
					CT314	Electrically powered	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem	27/10/2021	27/10/2021
									A070504	Premature Discharge of Battery	27/10/2021	27/10/2021
							F08	Hospitalization or Prolonged Hospitalization	A0705	Battery Problem	27/10/2021	27/10/2021
									A070504	Premature Discharge of Battery	27/10/2021	27/10/2021
							F1905	Device Revision or Replacement	A0705	Battery Problem	27/10/2021	27/10/2021
									A070504	Premature Discharge of Battery	27/10/2021	27/10/2021
					CT321	Long-term surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem	27/10/2021	27/10/2021
									A070504	Premature Discharge of Battery	27/10/2021	27/10/2021
							F08	Hospitalization or Prolonged Hospitalization	A0705	Battery Problem	27/10/2021	27/10/2021
									A070504	Premature Discharge of Battery	27/10/2021	27/10/2021
							F1905	Device Revision or Replacement	A0705	Battery Problem	27/10/2021	27/10/2021
									A070504	Premature Discharge of Battery	27/10/2021	27/10/2021
					CT449	Generators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem	27/10/2021	27/10/2021
									A070504	Premature Discharge of Battery	27/10/2021	27/10/2021
							F08	Hospitalization or Prolonged Hospitalization	A0705	Battery Problem	27/10/2021	27/10/2021
									A070504	Premature Discharge of Battery	27/10/2021	27/10/2021
							F1905	Device Revision or Replacement	A0705	Battery Problem	27/10/2021	27/10/2021
									A070504	Premature Discharge of Battery	27/10/2021	27/10/2021
					CT981	Single use	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem	27/10/2021	27/10/2021
									A070504	Premature Discharge of Battery	27/10/2021	27/10/2021
							F08	Hospitalization or Prolonged Hospitalization	A0705	Battery Problem	27/10/2021	27/10/2021
									A070504	Premature Discharge of Battery	27/10/2021	27/10/2021
							F1905	Device Revision or Replacement	A0705	Battery Problem	27/10/2021	27/10/2021
									A070504	Premature Discharge of Battery	27/10/2021	27/10/2021
					CT983	Surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem	27/10/2021	27/10/2021
									A070504	Premature Discharge of Battery	27/10/2021	27/10/2021
							F08	Hospitalization or Prolonged Hospitalization	A0705	Battery Problem	27/10/2021	27/10/2021
									A070504	Premature Discharge of Battery	27/10/2021	27/10/2021
							F1905	Device Revision or Replacement	A0705	Battery Problem	27/10/2021	27/10/2021
									A070504	Premature Discharge of Battery	27/10/2021	27/10/2021
6283	2021	All other reportable incidents	47265	Dual-chamber implantable pacemaker, rate-responsive	CT141	Electrodes/Leads and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070909	Oversensing		
							F26	No Health Consequences or Impact	A070909	Oversensing		
					CT145	Cardiovascular devices	E2403	No Clinical Signs, Symptoms or Conditions	A070909	Oversensing		
							F26	No Health Consequences or Impact	A070909	Oversensing		
					CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070909	Oversensing		
							F26	No Health Consequences or Impact	A070909	Oversensing		
					CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A070909	Oversensing		
							F26	No Health Consequences or Impact	A070909	Oversensing		
					CT233	Surgical	E2403	No Clinical Signs, Symptoms or Conditions	A070909	Oversensing		
							F26	No Health Consequences or Impact	A070909	Oversensing		
					CT2406	Implantable	E2403	No Clinical Signs, Symptoms or Conditions	A070909	Oversensing		
							F26	No Health Consequences or Impact	A070909	Oversensing		
					CT2407	Active implantable	E2403	No Clinical Signs, Symptoms or Conditions	A070909	Oversensing		
							F26	No Health Consequences or Impact	A070909	Oversensing		
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A070909	Oversensing		
							F26	No Health Consequences or Impact	A070909	Oversensing		
					CT2727	Active implant devices/ancillaries	E2403	No Clinical Signs, Symptoms or Conditions	A070909	Oversensing		
							F26	No Health Consequences or Impact	A070909	Oversensing		
					CT2750	Implantable cardiac electrical stimulators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070909	Oversensing		
							F26	No Health Consequences or Impact	A070909	Oversensing		
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A070909	Oversensing		
							F26	No Health Consequences or Impact	A070909	Oversensing		
					CT2985	Active	E2403	No Clinical Signs, Symptoms or Conditions	A070909	Oversensing		
							F26	No Health Consequences or Impact	A070909	Oversensing		
					CT3001	Implantable metal	E2403	No Clinical Signs, Symptoms or Conditions	A070909	Oversensing		
							F26	No Health Consequences or Impact	A070909	Oversensing		
					CT3002	Implantable synthetic polymer	E2403	No Clinical Signs, Symptoms or Conditions	A070909	Oversensing		
							F26	No Health Consequences or Impact	A070909	Oversensing		
					CT3006	Implantable inorganic material	E2403	No Clinical Signs, Symptoms or Conditions	A070909	Oversensing		
							F26	No Health Consequences or Impact	A070909	Oversensing		
					CT314	Electrically powered	E2403	No Clinical Signs, Symptoms or Conditions	A070909	Oversensing		
							F26	No Health Consequences or Impact	A070909	Oversensing		
					CT321	Long-term surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A070909	Oversensing		
							F26	No Health Consequences or Impact	A070909	Oversensing		
					CT449	Generators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070909	Oversensing		
							F26	No Health Consequences or Impact	A070909	Oversensing		
					CT981	Single use	E2403	No Clinical Signs, Symptoms or Conditions	A070909	Oversensing		
							F26	No Health Consequences or Impact	A070909	Oversensing		
					CT983	Surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A070909	Oversensing		
							F26	No Health Consequences or Impact	A070909	Oversensing		
6284	2021	All other reportable incidents	47265	Dual-chamber implantable pacemaker, rate-responsive	CT141	Electrodes/Leads and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	19/09/2022	19/09/2022
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	19/09/2022	19/09/2022
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	19/09/2022	19/09/2022
					CT145	Cardiovascular devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	19/09/2022	19/09/2022
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	19/09/2022	19/09/2022
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	19/09/2022	19/09/2022
					CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	19/09/2022	19/09/2022
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	19/09/2022	19/09/2022
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	19/09/2022	19/09/2022
					CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	19/09/2022	19/09/2022
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	19/09/2022	19/09/2022
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	19/09/2022	19/09/2022
					CT233	Surgical	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	19/09/2022	19/09/2022
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	19/09/2022	19/09/2022
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	19/09/2022	19/09/2022
					CT2406	Implantable	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	19/09/2022	19/09/2022
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	19/09/2022	19/09/2022

							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	19/09/2022	19/09/2022
					CT2407	Active implantable	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	19/09/2022	19/09/2022
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	19/09/2022	19/09/2022
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	19/09/2022	19/09/2022
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	19/09/2022	19/09/2022
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	19/09/2022	19/09/2022
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	19/09/2022	19/09/2022
					CT2727	Active implant devices/ancillaries	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	19/09/2022	19/09/2022
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	19/09/2022	19/09/2022
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	19/09/2022	19/09/2022
					CT2750	Implantable cardiac electrical stimulators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	19/09/2022	19/09/2022
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	19/09/2022	19/09/2022
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	19/09/2022	19/09/2022
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	19/09/2022	19/09/2022
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	19/09/2022	19/09/2022
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	19/09/2022	19/09/2022
					CT2985	Active	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	19/09/2022	19/09/2022
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	19/09/2022	19/09/2022
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	19/09/2022	19/09/2022
					CT3001	Implantable metal	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	19/09/2022	19/09/2022
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	19/09/2022	19/09/2022
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	19/09/2022	19/09/2022
					CT3002	Implantable synthetic polymer	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	19/09/2022	19/09/2022
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	19/09/2022	19/09/2022
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	19/09/2022	19/09/2022
					CT3006	Implantable inorganic material	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	19/09/2022	19/09/2022
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	19/09/2022	19/09/2022
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	19/09/2022	19/09/2022
					CT314	Electrically powered	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	19/09/2022	19/09/2022
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	19/09/2022	19/09/2022
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	19/09/2022	19/09/2022
					CT321	Long-term surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	19/09/2022	19/09/2022
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	19/09/2022	19/09/2022
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	19/09/2022	19/09/2022
					CT449	Generators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	19/09/2022	19/09/2022
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	19/09/2022	19/09/2022
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	19/09/2022	19/09/2022
					CT981	Single use	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	19/09/2022	19/09/2022
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	19/09/2022	19/09/2022
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	19/09/2022	19/09/2022
					CT983	Surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	19/09/2022	19/09/2022
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	19/09/2022	19/09/2022
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	19/09/2022	19/09/2022
6285	2021	All other reportable incidents	17882	Physiologic-monitoring defibrillation system	CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070803	Failure to Power Up		
							F27	Problem Identified During Non-Clinical Procedure	A070803	Failure to Power Up		
6286	2022	All other reportable incidents	47265	Dual-chamber implantable pacemaker, rate-responsive	CT141	Electrodes/Leads and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070101	Failure to Capture		
									A070102	High Capture Threshold		
							F26	No Health Consequences or Impact	A070101	Failure to Capture		
									A070102	High Capture Threshold		
					CT145	Cardiovascular devices	E2403	No Clinical Signs, Symptoms or Conditions	A070101	Failure to Capture		
									A070102	High Capture Threshold		
							F26	No Health Consequences or Impact	A070101	Failure to Capture		
									A070102	High Capture Threshold		
					CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070101	Failure to Capture		
									A070102	High Capture Threshold		
							F26	No Health Consequences or Impact	A070101	Failure to Capture		
									A070102	High Capture Threshold		
					CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A070101	Failure to Capture		
									A070102	High Capture Threshold		
							F26	No Health Consequences or Impact	A070101	Failure to Capture		
									A070102	High Capture Threshold		
					CT233	Surgical	E2403	No Clinical Signs, Symptoms or Conditions	A070101	Failure to Capture		
									A070102	High Capture Threshold		
							F26	No Health Consequences or Impact	A070101	Failure to Capture		
									A070102	High Capture Threshold		
					CT2406	Implantable	E2403	No Clinical Signs, Symptoms or Conditions	A070101	Failure to Capture		
									A070102	High Capture Threshold		
							F26	No Health Consequences or Impact	A070101	Failure to Capture		
									A070102	High Capture Threshold		
					CT2407	Active implantable	E2403	No Clinical Signs, Symptoms or Conditions	A070101	Failure to Capture		
									A070102	High Capture Threshold		
							F26	No Health Consequences or Impact	A070101	Failure to Capture		
									A070102	High Capture Threshold		
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A070101	Failure to Capture		
									A070102	High Capture Threshold		
							F26	No Health Consequences or Impact	A070101	Failure to Capture		
									A070102	High Capture Threshold		
					CT2727	Active implant devices/ancillaries	E2403	No Clinical Signs, Symptoms or Conditions	A070101	Failure to Capture		
									A070102	High Capture Threshold		
							F26	No Health Consequences or Impact	A070101	Failure to Capture		
									A070102	High Capture Threshold		
					CT2750	Implantable cardiac electrical stimulators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070101	Failure to Capture		
									A070102	High Capture Threshold		
							F26	No Health Consequences or Impact	A070101	Failure to Capture		
									A070102	High Capture Threshold		
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A070101	Failure to Capture		
									A070102	High Capture Threshold		
							F26	No Health Consequences or Impact	A070101	Failure to Capture		

									A070102	High Capture Threshold		
					CT2985	Active	E2403	No Clinical Signs, Symptoms or Conditions	A070101	Failure to Capture		
									A070102	High Capture Threshold		
							F26	No Health Consequences or Impact	A070101	Failure to Capture		
									A070102	High Capture Threshold		
					CT3001	Implantable metal	E2403	No Clinical Signs, Symptoms or Conditions	A070101	Failure to Capture		
									A070102	High Capture Threshold		
							F26	No Health Consequences or Impact	A070101	Failure to Capture		
									A070102	High Capture Threshold		
					CT3002	Implantable synthetic polymer	E2403	No Clinical Signs, Symptoms or Conditions	A070101	Failure to Capture		
									A070102	High Capture Threshold		
							F26	No Health Consequences or Impact	A070101	Failure to Capture		
									A070102	High Capture Threshold		
					CT3006	Implantable inorganic material	E2403	No Clinical Signs, Symptoms or Conditions	A070101	Failure to Capture		
									A070102	High Capture Threshold		
							F26	No Health Consequences or Impact	A070101	Failure to Capture		
									A070102	High Capture Threshold		
					CT314	Electrically powered	E2403	No Clinical Signs, Symptoms or Conditions	A070101	Failure to Capture		
									A070102	High Capture Threshold		
							F26	No Health Consequences or Impact	A070101	Failure to Capture		
									A070102	High Capture Threshold		
					CT321	Long-term surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A070101	Failure to Capture		
									A070102	High Capture Threshold		
							F26	No Health Consequences or Impact	A070101	Failure to Capture		
									A070102	High Capture Threshold		
					CT449	Generators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070101	Failure to Capture		
									A070102	High Capture Threshold		
							F26	No Health Consequences or Impact	A070101	Failure to Capture		
									A070102	High Capture Threshold		
					CT981	Single use	E2403	No Clinical Signs, Symptoms or Conditions	A070101	Failure to Capture		
									A070102	High Capture Threshold		
							F26	No Health Consequences or Impact	A070101	Failure to Capture		
									A070102	High Capture Threshold		
					CT983	Surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A070101	Failure to Capture		
									A070102	High Capture Threshold		
							F26	No Health Consequences or Impact	A070101	Failure to Capture		
									A070102	High Capture Threshold		
6287	2022	All other reportable incidents	17882	Physiologic-monitoring defibrillation system	CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A0719	Unexpected Shutdown		
							F27	Problem Identified During Non-Clinical Procedure	A0719	Unexpected Shutdown		
6288	2022	All other reportable incidents	17882	Physiologic-monitoring defibrillation system	CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A0719	Unexpected Shutdown		
							F26	No Health Consequences or Impact	A0719	Unexpected Shutdown		
6289	2022	All other reportable incidents	35223	Endocardial/interventricular septal pacing lead	CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070909	Oversensing		
									A090801	Signal Artifact/Noise		
							F26	No Health Consequences or Impact	A070909	Oversensing		
									A090801	Signal Artifact/Noise		
6290	2022	All other reportable incidents	35223	Endocardial/interventricular septal pacing lead	CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070909	Oversensing		
									A090801	Signal Artifact/Noise		
							F26	No Health Consequences or Impact	A070909	Oversensing		
									A090801	Signal Artifact/Noise		
6291	2022	All other reportable incidents	35852	Single-chamber implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem	01/04/2022	01/04/2022
									A070504	Premature Discharge of Battery	01/04/2022	01/04/2022
							F08	Hospitalization or Prolonged Hospitalization	A0705	Battery Problem	01/04/2022	01/04/2022
									A070504	Premature Discharge of Battery	01/04/2022	01/04/2022
							F1905	Device Revision or Replacement	A0705	Battery Problem	01/04/2022	01/04/2022
									A070504	Premature Discharge of Battery	01/04/2022	01/04/2022
6292	2022	All other reportable incidents	35852	Single-chamber implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem	31/05/2022	31/05/2022
									A070504	Premature Discharge of Battery	31/05/2022	31/05/2022
							F08	Hospitalization or Prolonged Hospitalization	A0705	Battery Problem	31/05/2022	31/05/2022
									A070504	Premature Discharge of Battery	31/05/2022	31/05/2022
							F1905	Device Revision or Replacement	A0705	Battery Problem	31/05/2022	31/05/2022
									A070504	Premature Discharge of Battery	31/05/2022	31/05/2022
6293	2021	All other reportable incidents	35852	Single-chamber implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem	21/09/2021	21/09/2021
									A070504	Premature Discharge of Battery	21/09/2021	21/09/2021
							F08	Hospitalization or Prolonged Hospitalization	A0705	Battery Problem	21/09/2021	21/09/2021
									A070504	Premature Discharge of Battery	21/09/2021	21/09/2021
							F1905	Device Revision or Replacement	A0705	Battery Problem	21/09/2021	21/09/2021
									A070504	Premature Discharge of Battery	21/09/2021	21/09/2021
6294	2022	All other reportable incidents	17882	Physiologic-monitoring defibrillation system	CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A071301	Failure to Deliver Shock/Stimulation		
							F27	Problem Identified During Non-Clinical Procedure	A071301	Failure to Deliver Shock/Stimulation		
6295	2021	All other reportable incidents	47267	Single-chamber implantable pacemaker, rate-responsive	CT141	Electrodes/Leads and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	27/07/2021	27/07/2021
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	27/07/2021	27/07/2021
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	27/07/2021	27/07/2021
					CT145	Cardiovascular devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	27/07/2021	27/07/2021
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	27/07/2021	27/07/2021
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	27/07/2021	27/07/2021
					CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	27/07/2021	27/07/2021
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	27/07/2021	27/07/2021
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	27/07/2021	27/07/2021
					CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	27/07/2021	27/07/2021
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	27/07/2021	27/07/2021
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	27/07/2021	27/07/2021
					CT233	Surgical	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	27/07/2021	27/07/2021
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	27/07/2021	27/07/2021
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	27/07/2021	27/07/2021
					CT2406	Implantable	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	27/07/2021	27/07/2021
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	27/07/2021	27/07/2021
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	27/07/2021	27/07/2021
					CT2407	Active implantable	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	27/07/2021	27/07/2021

							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	27/07/2021	27/07/2021
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	27/07/2021	27/07/2021
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	27/07/2021	27/07/2021
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	27/07/2021	27/07/2021
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	27/07/2021	27/07/2021
					CT2727	Active implant devices/ancillaries	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	27/07/2021	27/07/2021
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	27/07/2021	27/07/2021
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	27/07/2021	27/07/2021
					CT2750	Implantable cardiac electrical stimulators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	27/07/2021	27/07/2021
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	27/07/2021	27/07/2021
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	27/07/2021	27/07/2021
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	27/07/2021	27/07/2021
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	27/07/2021	27/07/2021
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	27/07/2021	27/07/2021
					CT2985	Active	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	27/07/2021	27/07/2021
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	27/07/2021	27/07/2021
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	27/07/2021	27/07/2021
					CT3001	Implantable metal	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	27/07/2021	27/07/2021
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	27/07/2021	27/07/2021
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	27/07/2021	27/07/2021
					CT3002	Implantable synthetic polymer	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	27/07/2021	27/07/2021
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	27/07/2021	27/07/2021
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	27/07/2021	27/07/2021
					CT3006	Implantable inorganic material	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	27/07/2021	27/07/2021
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	27/07/2021	27/07/2021
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	27/07/2021	27/07/2021
					CT314	Electrically powered	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	27/07/2021	27/07/2021
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	27/07/2021	27/07/2021
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	27/07/2021	27/07/2021
					CT321	Long-term surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	27/07/2021	27/07/2021
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	27/07/2021	27/07/2021
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	27/07/2021	27/07/2021
					CT449	Generators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	27/07/2021	27/07/2021
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	27/07/2021	27/07/2021
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	27/07/2021	27/07/2021
					CT981	Single use	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	27/07/2021	27/07/2021
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	27/07/2021	27/07/2021
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	27/07/2021	27/07/2021
					CT983	Surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	27/07/2021	27/07/2021
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	27/07/2021	27/07/2021
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	27/07/2021	27/07/2021
6296	2021	All other reportable incidents	47263	Cardiac resynchronization therapy implantable pacemaker	CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
6297	2021	All other reportable incidents	47265	Dual-chamber implantable pacemaker, rate-responsive	CT141	Electrodes/Leads and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	05/08/2021	05/08/2021
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	05/08/2021	05/08/2021
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	05/08/2021	05/08/2021
					CT145	Cardiovascular devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	05/08/2021	05/08/2021
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	05/08/2021	05/08/2021
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	05/08/2021	05/08/2021
					CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	05/08/2021	05/08/2021
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	05/08/2021	05/08/2021
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	05/08/2021	05/08/2021
					CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	05/08/2021	05/08/2021
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	05/08/2021	05/08/2021
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	05/08/2021	05/08/2021
					CT233	Surgical	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	05/08/2021	05/08/2021
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	05/08/2021	05/08/2021
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	05/08/2021	05/08/2021
					CT2406	Implantable	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	05/08/2021	05/08/2021
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	05/08/2021	05/08/2021
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	05/08/2021	05/08/2021
					CT2407	Active implantable	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	05/08/2021	05/08/2021
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	05/08/2021	05/08/2021
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	05/08/2021	05/08/2021
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	05/08/2021	05/08/2021
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	05/08/2021	05/08/2021
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	05/08/2021	05/08/2021
					CT2727	Active implant devices/ancillaries	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	05/08/2021	05/08/2021
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	05/08/2021	05/08/2021
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	05/08/2021	05/08/2021
					CT2750	Implantable cardiac electrical stimulators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	05/08/2021	05/08/2021
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	05/08/2021	05/08/2021
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	05/08/2021	05/08/2021
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	05/08/2021	05/08/2021
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	05/08/2021	05/08/2021
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	05/08/2021	05/08/2021
					CT2985	Active	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	05/08/2021	05/08/2021
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	05/08/2021	05/08/2021
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	05/08/2021	05/08/2021
					CT3001	Implantable metal	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	05/08/2021	05/08/2021
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	05/08/2021	05/08/2021
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	05/08/2021	05/08/2021

					CT3002	Implantable synthetic polymer	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	05/08/2021	05/08/2021
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	05/08/2021	05/08/2021
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	05/08/2021	05/08/2021
					CT3006	Implantable inorganic material	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	05/08/2021	05/08/2021
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	05/08/2021	05/08/2021
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	05/08/2021	05/08/2021
					CT314	Electrically powered	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	05/08/2021	05/08/2021
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	05/08/2021	05/08/2021
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	05/08/2021	05/08/2021
					CT321	Long-term surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	05/08/2021	05/08/2021
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	05/08/2021	05/08/2021
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	05/08/2021	05/08/2021
					CT449	Generators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	05/08/2021	05/08/2021
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	05/08/2021	05/08/2021
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	05/08/2021	05/08/2021
					CT981	Single use	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	05/08/2021	05/08/2021
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	05/08/2021	05/08/2021
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	05/08/2021	05/08/2021
					CT983	Surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	05/08/2021	05/08/2021
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	05/08/2021	05/08/2021
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	05/08/2021	05/08/2021
6298	2021	All other reportable incidents	47265	Dual-chamber implantable pacemaker, rate-responsive	CT141	Electrodes/Leads and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem	29/09/2021	29/09/2021
									A070504	Premature Discharge of Battery	29/09/2021	29/09/2021
							F08	Hospitalization or Prolonged Hospitalization	A0705	Battery Problem	29/09/2021	29/09/2021
									A070504	Premature Discharge of Battery	29/09/2021	29/09/2021
							F1905	Device Revision or Replacement	A0705	Battery Problem	29/09/2021	29/09/2021
									A070504	Premature Discharge of Battery	29/09/2021	29/09/2021
					CT145	Cardiovascular devices	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem	29/09/2021	29/09/2021
									A070504	Premature Discharge of Battery	29/09/2021	29/09/2021
							F08	Hospitalization or Prolonged Hospitalization	A0705	Battery Problem	29/09/2021	29/09/2021
									A070504	Premature Discharge of Battery	29/09/2021	29/09/2021
							F1905	Device Revision or Replacement	A0705	Battery Problem	29/09/2021	29/09/2021
									A070504	Premature Discharge of Battery	29/09/2021	29/09/2021
					CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem	29/09/2021	29/09/2021
									A070504	Premature Discharge of Battery	29/09/2021	29/09/2021
							F08	Hospitalization or Prolonged Hospitalization	A0705	Battery Problem	29/09/2021	29/09/2021
									A070504	Premature Discharge of Battery	29/09/2021	29/09/2021
							F1905	Device Revision or Replacement	A0705	Battery Problem	29/09/2021	29/09/2021
									A070504	Premature Discharge of Battery	29/09/2021	29/09/2021
					CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem	29/09/2021	29/09/2021
									A070504	Premature Discharge of Battery	29/09/2021	29/09/2021
							F08	Hospitalization or Prolonged Hospitalization	A0705	Battery Problem	29/09/2021	29/09/2021
									A070504	Premature Discharge of Battery	29/09/2021	29/09/2021
							F1905	Device Revision or Replacement	A0705	Battery Problem	29/09/2021	29/09/2021
									A070504	Premature Discharge of Battery	29/09/2021	29/09/2021
					CT233	Surgical	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem	29/09/2021	29/09/2021
									A070504	Premature Discharge of Battery	29/09/2021	29/09/2021
							F08	Hospitalization or Prolonged Hospitalization	A0705	Battery Problem	29/09/2021	29/09/2021
									A070504	Premature Discharge of Battery	29/09/2021	29/09/2021
							F1905	Device Revision or Replacement	A0705	Battery Problem	29/09/2021	29/09/2021
									A070504	Premature Discharge of Battery	29/09/2021	29/09/2021
					CT2406	Implantable	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem	29/09/2021	29/09/2021
									A070504	Premature Discharge of Battery	29/09/2021	29/09/2021
							F08	Hospitalization or Prolonged Hospitalization	A0705	Battery Problem	29/09/2021	29/09/2021
									A070504	Premature Discharge of Battery	29/09/2021	29/09/2021
							F1905	Device Revision or Replacement	A0705	Battery Problem	29/09/2021	29/09/2021
									A070504	Premature Discharge of Battery	29/09/2021	29/09/2021
					CT2407	Active implantable	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem	29/09/2021	29/09/2021
									A070504	Premature Discharge of Battery	29/09/2021	29/09/2021
							F08	Hospitalization or Prolonged Hospitalization	A0705	Battery Problem	29/09/2021	29/09/2021
									A070504	Premature Discharge of Battery	29/09/2021	29/09/2021
							F1905	Device Revision or Replacement	A0705	Battery Problem	29/09/2021	29/09/2021
									A070504	Premature Discharge of Battery	29/09/2021	29/09/2021
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem	29/09/2021	29/09/2021
									A070504	Premature Discharge of Battery	29/09/2021	29/09/2021
							F08	Hospitalization or Prolonged Hospitalization	A0705	Battery Problem	29/09/2021	29/09/2021
									A070504	Premature Discharge of Battery	29/09/2021	29/09/2021
							F1905	Device Revision or Replacement	A0705	Battery Problem	29/09/2021	29/09/2021
									A070504	Premature Discharge of Battery	29/09/2021	29/09/2021
					CT2727	Active implant devices/ancillaries	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem	29/09/2021	29/09/2021
									A070504	Premature Discharge of Battery	29/09/2021	29/09/2021
							F08	Hospitalization or Prolonged Hospitalization	A0705	Battery Problem	29/09/2021	29/09/2021
									A070504	Premature Discharge of Battery	29/09/2021	29/09/2021
							F1905	Device Revision or Replacement	A0705	Battery Problem	29/09/2021	29/09/2021
									A070504	Premature Discharge of Battery	29/09/2021	29/09/2021
					CT2750	Implantable cardiac electrical stimulators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem	29/09/2021	29/09/2021
									A070504	Premature Discharge of Battery	29/09/2021	29/09/2021
							F08	Hospitalization or Prolonged Hospitalization	A0705	Battery Problem	29/09/2021	29/09/2021
									A070504	Premature Discharge of Battery	29/09/2021	29/09/2021
							F1905	Device Revision or Replacement	A0705	Battery Problem	29/09/2021	29/09/2021
									A070504	Premature Discharge of Battery	29/09/2021	29/09/2021
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem	29/09/2021	29/09/2021
									A070504	Premature Discharge of Battery	29/09/2021	29/09/2021
							F08	Hospitalization or Prolonged Hospitalization	A0705	Battery Problem	29/09/2021	29/09/2021
									A070504	Premature Discharge of Battery	29/09/2021	29/09/2021
							F1905	Device Revision or Replacement	A0705	Battery Problem	29/09/2021	29/09/2021
									A070504	Premature Discharge of Battery	29/09/2021	29/09/2021
					CT2985	Active	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem	29/09/2021	29/09/2021

								A070504	Premature Discharge of Battery	29/09/2021	29/09/2021
						F08	Hospitalization or Prolonged Hospitalization	A0705	Battery Problem	29/09/2021	29/09/2021
								A070504	Premature Discharge of Battery	29/09/2021	29/09/2021
						F1905	Device Revision or Replacement	A0705	Battery Problem	29/09/2021	29/09/2021
								A070504	Premature Discharge of Battery	29/09/2021	29/09/2021
					CT3001	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem	29/09/2021	29/09/2021
					Implantable metal			A070504	Premature Discharge of Battery	29/09/2021	29/09/2021
						F08	Hospitalization or Prolonged Hospitalization	A0705	Battery Problem	29/09/2021	29/09/2021
								A070504	Premature Discharge of Battery	29/09/2021	29/09/2021
						F1905	Device Revision or Replacement	A0705	Battery Problem	29/09/2021	29/09/2021
								A070504	Premature Discharge of Battery	29/09/2021	29/09/2021
					CT3002	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem	29/09/2021	29/09/2021
					Implantable synthetic polymer			A070504	Premature Discharge of Battery	29/09/2021	29/09/2021
						F08	Hospitalization or Prolonged Hospitalization	A0705	Battery Problem	29/09/2021	29/09/2021
								A070504	Premature Discharge of Battery	29/09/2021	29/09/2021
						F1905	Device Revision or Replacement	A0705	Battery Problem	29/09/2021	29/09/2021
								A070504	Premature Discharge of Battery	29/09/2021	29/09/2021
					CT3006	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem	29/09/2021	29/09/2021
					Implantable inorganic material			A070504	Premature Discharge of Battery	29/09/2021	29/09/2021
						F08	Hospitalization or Prolonged Hospitalization	A0705	Battery Problem	29/09/2021	29/09/2021
								A070504	Premature Discharge of Battery	29/09/2021	29/09/2021
						F1905	Device Revision or Replacement	A0705	Battery Problem	29/09/2021	29/09/2021
								A070504	Premature Discharge of Battery	29/09/2021	29/09/2021
					CT314	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem	29/09/2021	29/09/2021
					Electrically powered			A070504	Premature Discharge of Battery	29/09/2021	29/09/2021
						F08	Hospitalization or Prolonged Hospitalization	A0705	Battery Problem	29/09/2021	29/09/2021
								A070504	Premature Discharge of Battery	29/09/2021	29/09/2021
						F1905	Device Revision or Replacement	A0705	Battery Problem	29/09/2021	29/09/2021
								A070504	Premature Discharge of Battery	29/09/2021	29/09/2021
					CT321	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem	29/09/2021	29/09/2021
					Long-term surgical invasive			A070504	Premature Discharge of Battery	29/09/2021	29/09/2021
						F08	Hospitalization or Prolonged Hospitalization	A0705	Battery Problem	29/09/2021	29/09/2021
								A070504	Premature Discharge of Battery	29/09/2021	29/09/2021
						F1905	Device Revision or Replacement	A0705	Battery Problem	29/09/2021	29/09/2021
								A070504	Premature Discharge of Battery	29/09/2021	29/09/2021
					CT449	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem	29/09/2021	29/09/2021
					Generators and associated devices			A070504	Premature Discharge of Battery	29/09/2021	29/09/2021
						F08	Hospitalization or Prolonged Hospitalization	A0705	Battery Problem	29/09/2021	29/09/2021
								A070504	Premature Discharge of Battery	29/09/2021	29/09/2021
						F1905	Device Revision or Replacement	A0705	Battery Problem	29/09/2021	29/09/2021
								A070504	Premature Discharge of Battery	29/09/2021	29/09/2021
					CT981	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem	29/09/2021	29/09/2021
					Single use			A070504	Premature Discharge of Battery	29/09/2021	29/09/2021
						F08	Hospitalization or Prolonged Hospitalization	A0705	Battery Problem	29/09/2021	29/09/2021
								A070504	Premature Discharge of Battery	29/09/2021	29/09/2021
						F1905	Device Revision or Replacement	A0705	Battery Problem	29/09/2021	29/09/2021
								A070504	Premature Discharge of Battery	29/09/2021	29/09/2021
					CT983	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem	29/09/2021	29/09/2021
					Surgical invasive			A070504	Premature Discharge of Battery	29/09/2021	29/09/2021
						F08	Hospitalization or Prolonged Hospitalization	A0705	Battery Problem	29/09/2021	29/09/2021
								A070504	Premature Discharge of Battery	29/09/2021	29/09/2021
						F1905	Device Revision or Replacement	A0705	Battery Problem	29/09/2021	29/09/2021
								A070504	Premature Discharge of Battery	29/09/2021	29/09/2021
6299	2021	All other reportable incidents	47263	Cardiac resynchronization therapy implantable pacemaker	CT1684	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem	23/09/2021	23/09/2021
					Cardiac pacemakers and associated devices			A070504	Premature Discharge of Battery	23/09/2021	23/09/2021
						F08	Hospitalization or Prolonged Hospitalization	A0705	Battery Problem	23/09/2021	23/09/2021
								A070504	Premature Discharge of Battery	23/09/2021	23/09/2021
						F1905	Device Revision or Replacement	A0705	Battery Problem	23/09/2021	23/09/2021
								A070504	Premature Discharge of Battery	23/09/2021	23/09/2021
					CT1685	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem	23/09/2021	23/09/2021
					Cardiac pacemakers			A070504	Premature Discharge of Battery	23/09/2021	23/09/2021
						F08	Hospitalization or Prolonged Hospitalization	A0705	Battery Problem	23/09/2021	23/09/2021
								A070504	Premature Discharge of Battery	23/09/2021	23/09/2021
						F1905	Device Revision or Replacement	A0705	Battery Problem	23/09/2021	23/09/2021
								A070504	Premature Discharge of Battery	23/09/2021	23/09/2021
					CT257	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem	23/09/2021	23/09/2021
					Implantable cardiac pacemakers			A070504	Premature Discharge of Battery	23/09/2021	23/09/2021
						F08	Hospitalization or Prolonged Hospitalization	A0705	Battery Problem	23/09/2021	23/09/2021
								A070504	Premature Discharge of Battery	23/09/2021	23/09/2021
						F1905	Device Revision or Replacement	A0705	Battery Problem	23/09/2021	23/09/2021
								A070504	Premature Discharge of Battery	23/09/2021	23/09/2021
					CT2866	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem	23/09/2021	23/09/2021
					Implantable cardiac electrical stimulators			A070504	Premature Discharge of Battery	23/09/2021	23/09/2021
						F08	Hospitalization or Prolonged Hospitalization	A0705	Battery Problem	23/09/2021	23/09/2021
								A070504	Premature Discharge of Battery	23/09/2021	23/09/2021
						F1905	Device Revision or Replacement	A0705	Battery Problem	23/09/2021	23/09/2021
								A070504	Premature Discharge of Battery	23/09/2021	23/09/2021
6300	2022	All other reportable incidents	47265	Dual-chamber implantable pacemaker, rate-responsive	CT141	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem	05/05/2022	05/05/2022
					Electrodes/Leads and associated devices			A070504	Premature Discharge of Battery	05/05/2022	05/05/2022
						F08	Hospitalization or Prolonged Hospitalization	A0705	Battery Problem	05/05/2022	05/05/2022
								A070504	Premature Discharge of Battery	05/05/2022	05/05/2022
						F1905	Device Revision or Replacement	A0705	Battery Problem	05/05/2022	05/05/2022
								A070504	Premature Discharge of Battery	05/05/2022	05/05/2022
					CT145	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem	05/05/2022	05/05/2022
					Cardiovascular devices			A070504	Premature Discharge of Battery	05/05/2022	05/05/2022
						F08	Hospitalization or Prolonged Hospitalization	A0705	Battery Problem	05/05/2022	05/05/2022
								A070504	Premature Discharge of Battery	05/05/2022	05/05/2022
						F1905	Device Revision or Replacement	A0705	Battery Problem	05/05/2022	05/05/2022

									A070504	Premature Discharge of Battery	05/05/2022	05/05/2022
					CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem	05/05/2022	05/05/2022
									A070504	Premature Discharge of Battery	05/05/2022	05/05/2022
							F08	Hospitalization or Prolonged Hospitalization	A0705	Battery Problem	05/05/2022	05/05/2022
									A070504	Premature Discharge of Battery	05/05/2022	05/05/2022
							F1905	Device Revision or Replacement	A0705	Battery Problem	05/05/2022	05/05/2022
									A070504	Premature Discharge of Battery	05/05/2022	05/05/2022
					CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem	05/05/2022	05/05/2022
									A070504	Premature Discharge of Battery	05/05/2022	05/05/2022
							F08	Hospitalization or Prolonged Hospitalization	A0705	Battery Problem	05/05/2022	05/05/2022
									A070504	Premature Discharge of Battery	05/05/2022	05/05/2022
							F1905	Device Revision or Replacement	A0705	Battery Problem	05/05/2022	05/05/2022
									A070504	Premature Discharge of Battery	05/05/2022	05/05/2022
					CT233	Surgical	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem	05/05/2022	05/05/2022
									A070504	Premature Discharge of Battery	05/05/2022	05/05/2022
							F08	Hospitalization or Prolonged Hospitalization	A0705	Battery Problem	05/05/2022	05/05/2022
									A070504	Premature Discharge of Battery	05/05/2022	05/05/2022
							F1905	Device Revision or Replacement	A0705	Battery Problem	05/05/2022	05/05/2022
									A070504	Premature Discharge of Battery	05/05/2022	05/05/2022
					CT2406	Implantable	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem	05/05/2022	05/05/2022
									A070504	Premature Discharge of Battery	05/05/2022	05/05/2022
							F08	Hospitalization or Prolonged Hospitalization	A0705	Battery Problem	05/05/2022	05/05/2022
									A070504	Premature Discharge of Battery	05/05/2022	05/05/2022
							F1905	Device Revision or Replacement	A0705	Battery Problem	05/05/2022	05/05/2022
									A070504	Premature Discharge of Battery	05/05/2022	05/05/2022
					CT2407	Active implantable	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem	05/05/2022	05/05/2022
									A070504	Premature Discharge of Battery	05/05/2022	05/05/2022
							F08	Hospitalization or Prolonged Hospitalization	A0705	Battery Problem	05/05/2022	05/05/2022
									A070504	Premature Discharge of Battery	05/05/2022	05/05/2022
							F1905	Device Revision or Replacement	A0705	Battery Problem	05/05/2022	05/05/2022
									A070504	Premature Discharge of Battery	05/05/2022	05/05/2022
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem	05/05/2022	05/05/2022
									A070504	Premature Discharge of Battery	05/05/2022	05/05/2022
							F08	Hospitalization or Prolonged Hospitalization	A0705	Battery Problem	05/05/2022	05/05/2022
									A070504	Premature Discharge of Battery	05/05/2022	05/05/2022
							F1905	Device Revision or Replacement	A0705	Battery Problem	05/05/2022	05/05/2022
									A070504	Premature Discharge of Battery	05/05/2022	05/05/2022
					CT2727	Active implant devices/ancillaries	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem	05/05/2022	05/05/2022
									A070504	Premature Discharge of Battery	05/05/2022	05/05/2022
							F08	Hospitalization or Prolonged Hospitalization	A0705	Battery Problem	05/05/2022	05/05/2022
									A070504	Premature Discharge of Battery	05/05/2022	05/05/2022
							F1905	Device Revision or Replacement	A0705	Battery Problem	05/05/2022	05/05/2022
									A070504	Premature Discharge of Battery	05/05/2022	05/05/2022
					CT2750	Implantable cardiac electrical stimulators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem	05/05/2022	05/05/2022
									A070504	Premature Discharge of Battery	05/05/2022	05/05/2022
							F08	Hospitalization or Prolonged Hospitalization	A0705	Battery Problem	05/05/2022	05/05/2022
									A070504	Premature Discharge of Battery	05/05/2022	05/05/2022
							F1905	Device Revision or Replacement	A0705	Battery Problem	05/05/2022	05/05/2022
									A070504	Premature Discharge of Battery	05/05/2022	05/05/2022
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem	05/05/2022	05/05/2022
									A070504	Premature Discharge of Battery	05/05/2022	05/05/2022
							F08	Hospitalization or Prolonged Hospitalization	A0705	Battery Problem	05/05/2022	05/05/2022
									A070504	Premature Discharge of Battery	05/05/2022	05/05/2022
							F1905	Device Revision or Replacement	A0705	Battery Problem	05/05/2022	05/05/2022
									A070504	Premature Discharge of Battery	05/05/2022	05/05/2022
					CT2985	Active	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem	05/05/2022	05/05/2022
									A070504	Premature Discharge of Battery	05/05/2022	05/05/2022
							F08	Hospitalization or Prolonged Hospitalization	A0705	Battery Problem	05/05/2022	05/05/2022
									A070504	Premature Discharge of Battery	05/05/2022	05/05/2022
							F1905	Device Revision or Replacement	A0705	Battery Problem	05/05/2022	05/05/2022
									A070504	Premature Discharge of Battery	05/05/2022	05/05/2022
					CT3001	Implantable metal	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem	05/05/2022	05/05/2022
									A070504	Premature Discharge of Battery	05/05/2022	05/05/2022
							F08	Hospitalization or Prolonged Hospitalization	A0705	Battery Problem	05/05/2022	05/05/2022
									A070504	Premature Discharge of Battery	05/05/2022	05/05/2022
							F1905	Device Revision or Replacement	A0705	Battery Problem	05/05/2022	05/05/2022
									A070504	Premature Discharge of Battery	05/05/2022	05/05/2022
					CT3002	Implantable synthetic polymer	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem	05/05/2022	05/05/2022
									A070504	Premature Discharge of Battery	05/05/2022	05/05/2022
							F08	Hospitalization or Prolonged Hospitalization	A0705	Battery Problem	05/05/2022	05/05/2022
									A070504	Premature Discharge of Battery	05/05/2022	05/05/2022
							F1905	Device Revision or Replacement	A0705	Battery Problem	05/05/2022	05/05/2022
									A070504	Premature Discharge of Battery	05/05/2022	05/05/2022
					CT3006	Implantable inorganic material	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem	05/05/2022	05/05/2022
									A070504	Premature Discharge of Battery	05/05/2022	05/05/2022
							F08	Hospitalization or Prolonged Hospitalization	A0705	Battery Problem	05/05/2022	05/05/2022
									A070504	Premature Discharge of Battery	05/05/2022	05/05/2022
							F1905	Device Revision or Replacement	A0705	Battery Problem	05/05/2022	05/05/2022
									A070504	Premature Discharge of Battery	05/05/2022	05/05/2022
					CT314	Electrically powered	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem	05/05/2022	05/05/2022
									A070504	Premature Discharge of Battery	05/05/2022	05/05/2022
							F08	Hospitalization or Prolonged Hospitalization	A0705	Battery Problem	05/05/2022	05/05/2022
									A070504	Premature Discharge of Battery	05/05/2022	05/05/2022
							F1905	Device Revision or Replacement	A0705	Battery Problem	05/05/2022	05/05/2022
									A070504	Premature Discharge of Battery	05/05/2022	05/05/2022
					CT321	Long-term surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem	05/05/2022	05/05/2022
									A070504	Premature Discharge of Battery	05/05/2022	05/05/2022
							F08	Hospitalization or Prolonged Hospitalization	A0705	Battery Problem	05/05/2022	05/05/2022

									A070504	Premature Discharge of Battery	05/05/2022	05/05/2022
							F1905	Device Revision or Replacement	A0705	Battery Problem	05/05/2022	05/05/2022
									A070504	Premature Discharge of Battery	05/05/2022	05/05/2022
					CT449	Generators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem	05/05/2022	05/05/2022
									A070504	Premature Discharge of Battery	05/05/2022	05/05/2022
							F08	Hospitalization or Prolonged Hospitalization	A0705	Battery Problem	05/05/2022	05/05/2022
									A070504	Premature Discharge of Battery	05/05/2022	05/05/2022
							F1905	Device Revision or Replacement	A0705	Battery Problem	05/05/2022	05/05/2022
									A070504	Premature Discharge of Battery	05/05/2022	05/05/2022
					CT981	Single use	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem	05/05/2022	05/05/2022
									A070504	Premature Discharge of Battery	05/05/2022	05/05/2022
							F08	Hospitalization or Prolonged Hospitalization	A0705	Battery Problem	05/05/2022	05/05/2022
									A070504	Premature Discharge of Battery	05/05/2022	05/05/2022
							F1905	Device Revision or Replacement	A0705	Battery Problem	05/05/2022	05/05/2022
									A070504	Premature Discharge of Battery	05/05/2022	05/05/2022
					CT983	Surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem	05/05/2022	05/05/2022
									A070504	Premature Discharge of Battery	05/05/2022	05/05/2022
							F08	Hospitalization or Prolonged Hospitalization	A0705	Battery Problem	05/05/2022	05/05/2022
									A070504	Premature Discharge of Battery	05/05/2022	05/05/2022
							F1905	Device Revision or Replacement	A0705	Battery Problem	05/05/2022	05/05/2022
									A070504	Premature Discharge of Battery	05/05/2022	05/05/2022
6301	2022	All other reportable incidents	36241	Pacemaker/defibrillator lead	CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A072201	High Impedance		
							F26	No Health Consequences or Impact	A072201	High Impedance		
6302	2022	All other reportable incidents	35852	Single-chamber implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
6303	2021	All other reportable incidents	37265	Dual-chamber implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	04/08/2021	04/08/2021
							F1901	Additional Surgery	A070504	Premature Discharge of Battery	04/08/2021	04/08/2021
6304	2022	All other reportable incidents	37265	Dual-chamber implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	24/03/2022	24/03/2022
									A071102	Failure to Interrogate	24/03/2022	24/03/2022
									A071206	Pacing Asynchronously	24/03/2022	24/03/2022
							F1901	Additional Surgery	A070504	Premature Discharge of Battery	24/03/2022	24/03/2022
									A071102	Failure to Interrogate	24/03/2022	24/03/2022
									A071206	Pacing Asynchronously	24/03/2022	24/03/2022
6305	2022	All other reportable incidents	47267	Single-chamber implantable pacemaker, rate-responsive	CT141	Electrodes/Leads and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	27/07/2022	27/07/2022
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	27/07/2022	27/07/2022
							F1903	Device Explantation	A070504	Premature Discharge of Battery	27/07/2022	27/07/2022
					CT145	Cardiovascular devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	27/07/2022	27/07/2022
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	27/07/2022	27/07/2022
							F1903	Device Explantation	A070504	Premature Discharge of Battery	27/07/2022	27/07/2022
					CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	27/07/2022	27/07/2022
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	27/07/2022	27/07/2022
							F1903	Device Explantation	A070504	Premature Discharge of Battery	27/07/2022	27/07/2022
					CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	27/07/2022	27/07/2022
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	27/07/2022	27/07/2022
							F1903	Device Explantation	A070504	Premature Discharge of Battery	27/07/2022	27/07/2022
					CT233	Surgical	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	27/07/2022	27/07/2022
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	27/07/2022	27/07/2022
							F1903	Device Explantation	A070504	Premature Discharge of Battery	27/07/2022	27/07/2022
					CT2406	Implantable	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	27/07/2022	27/07/2022
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	27/07/2022	27/07/2022
							F1903	Device Explantation	A070504	Premature Discharge of Battery	27/07/2022	27/07/2022
					CT2407	Active implantable	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	27/07/2022	27/07/2022
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	27/07/2022	27/07/2022
							F1903	Device Explantation	A070504	Premature Discharge of Battery	27/07/2022	27/07/2022
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	27/07/2022	27/07/2022
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	27/07/2022	27/07/2022
							F1903	Device Explantation	A070504	Premature Discharge of Battery	27/07/2022	27/07/2022
					CT2727	Active implant devices/ancillaries	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	27/07/2022	27/07/2022
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	27/07/2022	27/07/2022
							F1903	Device Explantation	A070504	Premature Discharge of Battery	27/07/2022	27/07/2022
					CT2750	Implantable cardiac electrical stimulators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	27/07/2022	27/07/2022
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	27/07/2022	27/07/2022
							F1903	Device Explantation	A070504	Premature Discharge of Battery	27/07/2022	27/07/2022
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	27/07/2022	27/07/2022
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	27/07/2022	27/07/2022
							F1903	Device Explantation	A070504	Premature Discharge of Battery	27/07/2022	27/07/2022
					CT2985	Active	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	27/07/2022	27/07/2022
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	27/07/2022	27/07/2022
							F1903	Device Explantation	A070504	Premature Discharge of Battery	27/07/2022	27/07/2022
					CT3001	Implantable metal	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	27/07/2022	27/07/2022
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	27/07/2022	27/07/2022
							F1903	Device Explantation	A070504	Premature Discharge of Battery	27/07/2022	27/07/2022
					CT3002	Implantable synthetic polymer	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	27/07/2022	27/07/2022
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	27/07/2022	27/07/2022
							F1903	Device Explantation	A070504	Premature Discharge of Battery	27/07/2022	27/07/2022
					CT3006	Implantable inorganic material	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	27/07/2022	27/07/2022
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	27/07/2022	27/07/2022
							F1903	Device Explantation	A070504	Premature Discharge of Battery	27/07/2022	27/07/2022
					CT314	Electrically powered	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	27/07/2022	27/07/2022
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	27/07/2022	27/07/2022
							F1903	Device Explantation	A070504	Premature Discharge of Battery	27/07/2022	27/07/2022
					CT321	Long-term surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	27/07/2022	27/07/2022
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	27/07/2022	27/07/2022
							F1903	Device Explantation	A070504	Premature Discharge of Battery	27/07/2022	27/07/2022
					CT449	Generators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	27/07/2022	27/07/2022
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	27/07/2022	27/07/2022
							F1903	Device Explantation	A070504	Premature Discharge of Battery	27/07/2022	27/07/2022
					CT981	Single use	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	27/07/2022	27/07/2022

						F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	27/07/2022	27/07/2022
						F1903	Device Explantation	A070504	Premature Discharge of Battery	27/07/2022	27/07/2022
				CT983	Surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	27/07/2022	27/07/2022
						F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	27/07/2022	27/07/2022
						F1903	Device Explantation	A070504	Premature Discharge of Battery	27/07/2022	27/07/2022
6306	2021	All other reportable incidents	47205	Cardiac pulse generator programmer	CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A13	Communication or Transmission Problem		
							F26	No Health Consequences or Impact	A13	Communication or Transmission Problem		
6307	2021	All other reportable incidents	47265	Dual-chamber implantable pacemaker, rate-responsive	CT141	Electrodes/Leads and associated devices	E0602	Cardiac Arrest	A0712	Pacing Problem		
							E2403	No Clinical Signs, Symptoms or Conditions	A0712	Pacing Problem		
							F1203	Life Threatening Illness or Injury	A0712	Pacing Problem		
					CT145	Cardiovascular devices	E0602	Cardiac Arrest	A0712	Pacing Problem		
							E2403	No Clinical Signs, Symptoms or Conditions	A0712	Pacing Problem		
							F1203	Life Threatening Illness or Injury	A0712	Pacing Problem		
					CT1684	Cardiac pacemakers and associated devices	E0602	Cardiac Arrest	A0712	Pacing Problem		
							E2403	No Clinical Signs, Symptoms or Conditions	A0712	Pacing Problem		
							F1203	Life Threatening Illness or Injury	A0712	Pacing Problem		
					CT1685	Cardiac pacemakers	E0602	Cardiac Arrest	A0712	Pacing Problem		
							E2403	No Clinical Signs, Symptoms or Conditions	A0712	Pacing Problem		
							F1203	Life Threatening Illness or Injury	A0712	Pacing Problem		
					CT233	Surgical	E0602	Cardiac Arrest	A0712	Pacing Problem		
							E2403	No Clinical Signs, Symptoms or Conditions	A0712	Pacing Problem		
							F1203	Life Threatening Illness or Injury	A0712	Pacing Problem		
					CT2406	Implantable	E0602	Cardiac Arrest	A0712	Pacing Problem		
							E2403	No Clinical Signs, Symptoms or Conditions	A0712	Pacing Problem		
							F1203	Life Threatening Illness or Injury	A0712	Pacing Problem		
					CT2407	Active implantable	E0602	Cardiac Arrest	A0712	Pacing Problem		
							E2403	No Clinical Signs, Symptoms or Conditions	A0712	Pacing Problem		
							F1203	Life Threatening Illness or Injury	A0712	Pacing Problem		
					CT257	Implantable cardiac pacemakers	E0602	Cardiac Arrest	A0712	Pacing Problem		
							E2403	No Clinical Signs, Symptoms or Conditions	A0712	Pacing Problem		
							F1203	Life Threatening Illness or Injury	A0712	Pacing Problem		
					CT2727	Active implant devices/ancillaries	E0602	Cardiac Arrest	A0712	Pacing Problem		
							E2403	No Clinical Signs, Symptoms or Conditions	A0712	Pacing Problem		
							F1203	Life Threatening Illness or Injury	A0712	Pacing Problem		
					CT2750	Implantable cardiac electrical stimulators and associated devices	E0602	Cardiac Arrest	A0712	Pacing Problem		
							E2403	No Clinical Signs, Symptoms or Conditions	A0712	Pacing Problem		
							F1203	Life Threatening Illness or Injury	A0712	Pacing Problem		
					CT2866	Implantable cardiac electrical stimulators	E0602	Cardiac Arrest	A0712	Pacing Problem		
							E2403	No Clinical Signs, Symptoms or Conditions	A0712	Pacing Problem		
							F1203	Life Threatening Illness or Injury	A0712	Pacing Problem		
					CT2985	Active	E0602	Cardiac Arrest	A0712	Pacing Problem		
							E2403	No Clinical Signs, Symptoms or Conditions	A0712	Pacing Problem		
							F1203	Life Threatening Illness or Injury	A0712	Pacing Problem		
					CT3001	Implantable metal	E0602	Cardiac Arrest	A0712	Pacing Problem		
							E2403	No Clinical Signs, Symptoms or Conditions	A0712	Pacing Problem		
							F1203	Life Threatening Illness or Injury	A0712	Pacing Problem		
					CT3002	Implantable synthetic polymer	E0602	Cardiac Arrest	A0712	Pacing Problem		
							E2403	No Clinical Signs, Symptoms or Conditions	A0712	Pacing Problem		
							F1203	Life Threatening Illness or Injury	A0712	Pacing Problem		
					CT3006	Implantable inorganic material	E0602	Cardiac Arrest	A0712	Pacing Problem		
							E2403	No Clinical Signs, Symptoms or Conditions	A0712	Pacing Problem		
							F1203	Life Threatening Illness or Injury	A0712	Pacing Problem		
					CT314	Electrically powered	E0602	Cardiac Arrest	A0712	Pacing Problem		
							E2403	No Clinical Signs, Symptoms or Conditions	A0712	Pacing Problem		
							F1203	Life Threatening Illness or Injury	A0712	Pacing Problem		
					CT321	Long-term surgical invasive	E0602	Cardiac Arrest	A0712	Pacing Problem		
							E2403	No Clinical Signs, Symptoms or Conditions	A0712	Pacing Problem		
							F1203	Life Threatening Illness or Injury	A0712	Pacing Problem		
					CT449	Generators and associated devices	E0602	Cardiac Arrest	A0712	Pacing Problem		
							E2403	No Clinical Signs, Symptoms or Conditions	A0712	Pacing Problem		
							F1203	Life Threatening Illness or Injury	A0712	Pacing Problem		
					CT981	Single use	E0602	Cardiac Arrest	A0712	Pacing Problem		
							E2403	No Clinical Signs, Symptoms or Conditions	A0712	Pacing Problem		
							F1203	Life Threatening Illness or Injury	A0712	Pacing Problem		
					CT983	Surgical invasive	E0602	Cardiac Arrest	A0712	Pacing Problem		
							E2403	No Clinical Signs, Symptoms or Conditions	A0712	Pacing Problem		
							F1203	Life Threatening Illness or Injury	A0712	Pacing Problem		
6308	2021	All other reportable incidents	35852	Single-chamber implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	23/09/2021	23/09/2021
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	23/09/2021	23/09/2021
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	23/09/2021	23/09/2021
6309	2021	All other reportable incidents	35852	Single-chamber implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	07/09/2021	07/09/2021
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	07/09/2021	07/09/2021
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	07/09/2021	07/09/2021
6310	2021	All other reportable incidents	37265	Dual-chamber implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	05/10/2021	05/10/2021
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	05/10/2021	05/10/2021
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	05/10/2021	05/10/2021
6311	2021	All other reportable incidents	47267	Single-chamber implantable pacemaker, rate-responsive	CT141	Electrodes/Leads and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	23/08/2021	23/08/2021
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	23/08/2021	23/08/2021
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	23/08/2021	23/08/2021
					CT145	Cardiovascular devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	23/08/2021	23/08/2021
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	23/08/2021	23/08/2021
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	23/08/2021	23/08/2021
					CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	23/08/2021	23/08/2021
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	23/08/2021	23/08/2021
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	23/08/2021	23/08/2021
					CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	23/08/2021	23/08/2021
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	23/08/2021	23/08/2021
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	23/08/2021	23/08/2021

					CT233	Surgical	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	23/08/2021	23/08/2021
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	23/08/2021	23/08/2021
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	23/08/2021	23/08/2021
					CT2406	Implantable	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	23/08/2021	23/08/2021
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	23/08/2021	23/08/2021
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	23/08/2021	23/08/2021
					CT2407	Active implantable	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	23/08/2021	23/08/2021
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	23/08/2021	23/08/2021
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	23/08/2021	23/08/2021
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	23/08/2021	23/08/2021
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	23/08/2021	23/08/2021
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	23/08/2021	23/08/2021
					CT2727	Active implant devices/ancillaries	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	23/08/2021	23/08/2021
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	23/08/2021	23/08/2021
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	23/08/2021	23/08/2021
					CT2750	Implantable cardiac electrical stimulators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	23/08/2021	23/08/2021
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	23/08/2021	23/08/2021
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	23/08/2021	23/08/2021
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	23/08/2021	23/08/2021
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	23/08/2021	23/08/2021
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	23/08/2021	23/08/2021
					CT2985	Active	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	23/08/2021	23/08/2021
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	23/08/2021	23/08/2021
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	23/08/2021	23/08/2021
					CT3001	Implantable metal	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	23/08/2021	23/08/2021
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	23/08/2021	23/08/2021
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	23/08/2021	23/08/2021
					CT3002	Implantable synthetic polymer	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	23/08/2021	23/08/2021
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	23/08/2021	23/08/2021
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	23/08/2021	23/08/2021
					CT3006	Implantable inorganic material	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	23/08/2021	23/08/2021
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	23/08/2021	23/08/2021
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	23/08/2021	23/08/2021
					CT314	Electrically powered	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	23/08/2021	23/08/2021
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	23/08/2021	23/08/2021
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	23/08/2021	23/08/2021
					CT321	Long-term surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	23/08/2021	23/08/2021
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	23/08/2021	23/08/2021
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	23/08/2021	23/08/2021
					CT449	Generators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	23/08/2021	23/08/2021
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	23/08/2021	23/08/2021
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	23/08/2021	23/08/2021
					CT981	Single use	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	23/08/2021	23/08/2021
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	23/08/2021	23/08/2021
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	23/08/2021	23/08/2021
					CT983	Surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	23/08/2021	23/08/2021
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	23/08/2021	23/08/2021
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	23/08/2021	23/08/2021
6312	2022	All other reportable incidents	47270	Cardiac resynchronization therapy implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	08/04/2022	08/04/2022
							F1901	Additional Surgery	A070504	Premature Discharge of Battery	08/04/2022	08/04/2022
6313	2022	All other reportable incidents	17882	Physiologic-monitoring defibrillation system	CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070802	Intermittent Loss of Power		
							F27	Problem Identified During Non-Clinical Procedure	A070802	Intermittent Loss of Power		
6314	2022	All other reportable incidents	47270	Cardiac resynchronization therapy implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery		
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
6315	2022	All other reportable incidents	47265	Dual-chamber implantable pacemaker, rate-responsive	CT141	Electrodes/Leads and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem	29/04/2022	29/04/2022
									A070504	Premature Discharge of Battery	29/04/2022	29/04/2022
							F08	Hospitalization or Prolonged Hospitalization	A0705	Battery Problem	29/04/2022	29/04/2022
									A070504	Premature Discharge of Battery	29/04/2022	29/04/2022
							F1905	Device Revision or Replacement	A0705	Battery Problem	29/04/2022	29/04/2022
									A070504	Premature Discharge of Battery	29/04/2022	29/04/2022
					CT145	Cardiovascular devices	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem	29/04/2022	29/04/2022
									A070504	Premature Discharge of Battery	29/04/2022	29/04/2022
							F08	Hospitalization or Prolonged Hospitalization	A0705	Battery Problem	29/04/2022	29/04/2022
									A070504	Premature Discharge of Battery	29/04/2022	29/04/2022
							F1905	Device Revision or Replacement	A0705	Battery Problem	29/04/2022	29/04/2022
									A070504	Premature Discharge of Battery	29/04/2022	29/04/2022
					CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem	29/04/2022	29/04/2022
									A070504	Premature Discharge of Battery	29/04/2022	29/04/2022
							F08	Hospitalization or Prolonged Hospitalization	A0705	Battery Problem	29/04/2022	29/04/2022
									A070504	Premature Discharge of Battery	29/04/2022	29/04/2022
							F1905	Device Revision or Replacement	A0705	Battery Problem	29/04/2022	29/04/2022
									A070504	Premature Discharge of Battery	29/04/2022	29/04/2022
					CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem	29/04/2022	29/04/2022
									A070504	Premature Discharge of Battery	29/04/2022	29/04/2022
							F08	Hospitalization or Prolonged Hospitalization	A0705	Battery Problem	29/04/2022	29/04/2022
									A070504	Premature Discharge of Battery	29/04/2022	29/04/2022
							F1905	Device Revision or Replacement	A0705	Battery Problem	29/04/2022	29/04/2022
									A070504	Premature Discharge of Battery	29/04/2022	29/04/2022
					CT233	Surgical	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem	29/04/2022	29/04/2022
									A070504	Premature Discharge of Battery	29/04/2022	29/04/2022
							F08	Hospitalization or Prolonged Hospitalization	A0705	Battery Problem	29/04/2022	29/04/2022
									A070504	Premature Discharge of Battery	29/04/2022	29/04/2022
							F1905	Device Revision or Replacement	A0705	Battery Problem	29/04/2022	29/04/2022
									A070504	Premature Discharge of Battery	29/04/2022	29/04/2022
					CT2406	Implantable	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem	29/04/2022	29/04/2022
									A070504	Premature Discharge of Battery	29/04/2022	29/04/2022

							F08	Hospitalization or Prolonged Hospitalization	A0705	Battery Problem	29/04/2022	29/04/2022
									A070504	Premature Discharge of Battery	29/04/2022	29/04/2022
							F1905	Device Revision or Replacement	A0705	Battery Problem	29/04/2022	29/04/2022
									A070504	Premature Discharge of Battery	29/04/2022	29/04/2022
					CT2407	Active implantable	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem	29/04/2022	29/04/2022
									A070504	Premature Discharge of Battery	29/04/2022	29/04/2022
							F08	Hospitalization or Prolonged Hospitalization	A0705	Battery Problem	29/04/2022	29/04/2022
									A070504	Premature Discharge of Battery	29/04/2022	29/04/2022
							F1905	Device Revision or Replacement	A0705	Battery Problem	29/04/2022	29/04/2022
									A070504	Premature Discharge of Battery	29/04/2022	29/04/2022
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem	29/04/2022	29/04/2022
									A070504	Premature Discharge of Battery	29/04/2022	29/04/2022
							F08	Hospitalization or Prolonged Hospitalization	A0705	Battery Problem	29/04/2022	29/04/2022
									A070504	Premature Discharge of Battery	29/04/2022	29/04/2022
							F1905	Device Revision or Replacement	A0705	Battery Problem	29/04/2022	29/04/2022
									A070504	Premature Discharge of Battery	29/04/2022	29/04/2022
					CT2727	Active implant devices/ancillaries	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem	29/04/2022	29/04/2022
									A070504	Premature Discharge of Battery	29/04/2022	29/04/2022
							F08	Hospitalization or Prolonged Hospitalization	A0705	Battery Problem	29/04/2022	29/04/2022
									A070504	Premature Discharge of Battery	29/04/2022	29/04/2022
							F1905	Device Revision or Replacement	A0705	Battery Problem	29/04/2022	29/04/2022
									A070504	Premature Discharge of Battery	29/04/2022	29/04/2022
					CT2750	Implantable cardiac electrical stimulators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem	29/04/2022	29/04/2022
									A070504	Premature Discharge of Battery	29/04/2022	29/04/2022
							F08	Hospitalization or Prolonged Hospitalization	A0705	Battery Problem	29/04/2022	29/04/2022
									A070504	Premature Discharge of Battery	29/04/2022	29/04/2022
							F1905	Device Revision or Replacement	A0705	Battery Problem	29/04/2022	29/04/2022
									A070504	Premature Discharge of Battery	29/04/2022	29/04/2022
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem	29/04/2022	29/04/2022
									A070504	Premature Discharge of Battery	29/04/2022	29/04/2022
							F08	Hospitalization or Prolonged Hospitalization	A0705	Battery Problem	29/04/2022	29/04/2022
									A070504	Premature Discharge of Battery	29/04/2022	29/04/2022
							F1905	Device Revision or Replacement	A0705	Battery Problem	29/04/2022	29/04/2022
									A070504	Premature Discharge of Battery	29/04/2022	29/04/2022
					CT2985	Active	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem	29/04/2022	29/04/2022
									A070504	Premature Discharge of Battery	29/04/2022	29/04/2022
							F08	Hospitalization or Prolonged Hospitalization	A0705	Battery Problem	29/04/2022	29/04/2022
									A070504	Premature Discharge of Battery	29/04/2022	29/04/2022
							F1905	Device Revision or Replacement	A0705	Battery Problem	29/04/2022	29/04/2022
									A070504	Premature Discharge of Battery	29/04/2022	29/04/2022
					CT3001	Implantable metal	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem	29/04/2022	29/04/2022
									A070504	Premature Discharge of Battery	29/04/2022	29/04/2022
							F08	Hospitalization or Prolonged Hospitalization	A0705	Battery Problem	29/04/2022	29/04/2022
									A070504	Premature Discharge of Battery	29/04/2022	29/04/2022
							F1905	Device Revision or Replacement	A0705	Battery Problem	29/04/2022	29/04/2022
									A070504	Premature Discharge of Battery	29/04/2022	29/04/2022
					CT3002	Implantable synthetic polymer	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem	29/04/2022	29/04/2022
									A070504	Premature Discharge of Battery	29/04/2022	29/04/2022
							F08	Hospitalization or Prolonged Hospitalization	A0705	Battery Problem	29/04/2022	29/04/2022
									A070504	Premature Discharge of Battery	29/04/2022	29/04/2022
							F1905	Device Revision or Replacement	A0705	Battery Problem	29/04/2022	29/04/2022
									A070504	Premature Discharge of Battery	29/04/2022	29/04/2022
					CT3006	Implantable inorganic material	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem	29/04/2022	29/04/2022
									A070504	Premature Discharge of Battery	29/04/2022	29/04/2022
							F08	Hospitalization or Prolonged Hospitalization	A0705	Battery Problem	29/04/2022	29/04/2022
									A070504	Premature Discharge of Battery	29/04/2022	29/04/2022
							F1905	Device Revision or Replacement	A0705	Battery Problem	29/04/2022	29/04/2022
									A070504	Premature Discharge of Battery	29/04/2022	29/04/2022
					CT314	Electrically powered	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem	29/04/2022	29/04/2022
									A070504	Premature Discharge of Battery	29/04/2022	29/04/2022
							F08	Hospitalization or Prolonged Hospitalization	A0705	Battery Problem	29/04/2022	29/04/2022
									A070504	Premature Discharge of Battery	29/04/2022	29/04/2022
							F1905	Device Revision or Replacement	A0705	Battery Problem	29/04/2022	29/04/2022
									A070504	Premature Discharge of Battery	29/04/2022	29/04/2022
					CT321	Long-term surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem	29/04/2022	29/04/2022
									A070504	Premature Discharge of Battery	29/04/2022	29/04/2022
							F08	Hospitalization or Prolonged Hospitalization	A0705	Battery Problem	29/04/2022	29/04/2022
									A070504	Premature Discharge of Battery	29/04/2022	29/04/2022
							F1905	Device Revision or Replacement	A0705	Battery Problem	29/04/2022	29/04/2022
									A070504	Premature Discharge of Battery	29/04/2022	29/04/2022
					CT449	Generators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem	29/04/2022	29/04/2022
									A070504	Premature Discharge of Battery	29/04/2022	29/04/2022
							F08	Hospitalization or Prolonged Hospitalization	A0705	Battery Problem	29/04/2022	29/04/2022
									A070504	Premature Discharge of Battery	29/04/2022	29/04/2022
							F1905	Device Revision or Replacement	A0705	Battery Problem	29/04/2022	29/04/2022
									A070504	Premature Discharge of Battery	29/04/2022	29/04/2022
					CT981	Single use	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem	29/04/2022	29/04/2022
									A070504	Premature Discharge of Battery	29/04/2022	29/04/2022
							F08	Hospitalization or Prolonged Hospitalization	A0705	Battery Problem	29/04/2022	29/04/2022
									A070504	Premature Discharge of Battery	29/04/2022	29/04/2022
							F1905	Device Revision or Replacement	A0705	Battery Problem	29/04/2022	29/04/2022
									A070504	Premature Discharge of Battery	29/04/2022	29/04/2022
					CT983	Surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem	29/04/2022	29/04/2022
									A070504	Premature Discharge of Battery	29/04/2022	29/04/2022
							F08	Hospitalization or Prolonged Hospitalization	A0705	Battery Problem	29/04/2022	29/04/2022
									A070504	Premature Discharge of Battery	29/04/2022	29/04/2022
							F1905	Device Revision or Replacement	A0705	Battery Problem	29/04/2022	29/04/2022
									A070504	Premature Discharge of Battery	29/04/2022	29/04/2022

6316	2022	All other reportable incidents	47267	Single-chamber implantable pacemaker, rate-responsive	CT141	Electrodes/Leads and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT145	Cardiovascular devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT233	Surgical	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT2406	Implantable	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT2407	Active implantable	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT2727	Active implant devices/ancillaries	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT2750	Implantable cardiac electrical stimulators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT2985	Active	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT3001	Implantable metal	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT3002	Implantable synthetic polymer	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT3006	Implantable inorganic material	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT314	Electrically powered	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT321	Long-term surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT449	Generators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT981	Single use	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT983	Surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
6317	2021	All other reportable incidents	47270	Cardiac resynchronization therapy implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	17/08/2021	17/08/2021
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery	17/08/2021	17/08/2021
6318	2022	All other reportable incidents	37265	Dual-chamber implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem		
							F26	No Health Consequences or Impact	A0705	Battery Problem		
6319	2021	All other reportable incidents	35852	Single-chamber implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	10/08/2021	10/08/2021
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	10/08/2021	10/08/2021
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	10/08/2021	10/08/2021
6320	2021	All other reportable incidents	47265	Dual-chamber implantable pacemaker, rate-responsive	CT141	Electrodes/Leads and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT145	Cardiovascular devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT233	Surgical	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT2406	Implantable	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT2407	Active implantable	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT2727	Active implant devices/ancillaries	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT2750	Implantable cardiac electrical stimulators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT2985	Active	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT3001	Implantable metal	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT3002	Implantable synthetic polymer	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT3006	Implantable inorganic material	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT314	Electrically powered	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT321	Long-term surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT449	Generators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT981	Single use	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT983	Surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
6321	2021	All other reportable incidents	47264	Dual-chamber implantable pacemaker, demand	CT141	Electrodes/Leads and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem	09/08/2021	09/08/2021

									A070504	Premature Discharge of Battery	09/08/2021	09/08/2021
							F08	Hospitalization or Prolonged Hospitalization	A0705	Battery Problem	09/08/2021	09/08/2021
									A070504	Premature Discharge of Battery	09/08/2021	09/08/2021
							F1905	Device Revision or Replacement	A0705	Battery Problem	09/08/2021	09/08/2021
									A070504	Premature Discharge of Battery	09/08/2021	09/08/2021
					CT145	Cardiovascular devices	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem	09/08/2021	09/08/2021
									A070504	Premature Discharge of Battery	09/08/2021	09/08/2021
							F08	Hospitalization or Prolonged Hospitalization	A0705	Battery Problem	09/08/2021	09/08/2021
									A070504	Premature Discharge of Battery	09/08/2021	09/08/2021
							F1905	Device Revision or Replacement	A0705	Battery Problem	09/08/2021	09/08/2021
									A070504	Premature Discharge of Battery	09/08/2021	09/08/2021
					CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem	09/08/2021	09/08/2021
									A070504	Premature Discharge of Battery	09/08/2021	09/08/2021
							F08	Hospitalization or Prolonged Hospitalization	A0705	Battery Problem	09/08/2021	09/08/2021
									A070504	Premature Discharge of Battery	09/08/2021	09/08/2021
							F1905	Device Revision or Replacement	A0705	Battery Problem	09/08/2021	09/08/2021
									A070504	Premature Discharge of Battery	09/08/2021	09/08/2021
					CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem	09/08/2021	09/08/2021
									A070504	Premature Discharge of Battery	09/08/2021	09/08/2021
							F08	Hospitalization or Prolonged Hospitalization	A0705	Battery Problem	09/08/2021	09/08/2021
									A070504	Premature Discharge of Battery	09/08/2021	09/08/2021
							F1905	Device Revision or Replacement	A0705	Battery Problem	09/08/2021	09/08/2021
									A070504	Premature Discharge of Battery	09/08/2021	09/08/2021
					CT233	Surgical	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem	09/08/2021	09/08/2021
									A070504	Premature Discharge of Battery	09/08/2021	09/08/2021
							F08	Hospitalization or Prolonged Hospitalization	A0705	Battery Problem	09/08/2021	09/08/2021
									A070504	Premature Discharge of Battery	09/08/2021	09/08/2021
							F1905	Device Revision or Replacement	A0705	Battery Problem	09/08/2021	09/08/2021
									A070504	Premature Discharge of Battery	09/08/2021	09/08/2021
					CT2406	Implantable	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem	09/08/2021	09/08/2021
									A070504	Premature Discharge of Battery	09/08/2021	09/08/2021
							F08	Hospitalization or Prolonged Hospitalization	A0705	Battery Problem	09/08/2021	09/08/2021
									A070504	Premature Discharge of Battery	09/08/2021	09/08/2021
							F1905	Device Revision or Replacement	A0705	Battery Problem	09/08/2021	09/08/2021
									A070504	Premature Discharge of Battery	09/08/2021	09/08/2021
					CT2407	Active implantable	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem	09/08/2021	09/08/2021
									A070504	Premature Discharge of Battery	09/08/2021	09/08/2021
							F08	Hospitalization or Prolonged Hospitalization	A0705	Battery Problem	09/08/2021	09/08/2021
									A070504	Premature Discharge of Battery	09/08/2021	09/08/2021
							F1905	Device Revision or Replacement	A0705	Battery Problem	09/08/2021	09/08/2021
									A070504	Premature Discharge of Battery	09/08/2021	09/08/2021
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem	09/08/2021	09/08/2021
									A070504	Premature Discharge of Battery	09/08/2021	09/08/2021
							F08	Hospitalization or Prolonged Hospitalization	A0705	Battery Problem	09/08/2021	09/08/2021
									A070504	Premature Discharge of Battery	09/08/2021	09/08/2021
							F1905	Device Revision or Replacement	A0705	Battery Problem	09/08/2021	09/08/2021
									A070504	Premature Discharge of Battery	09/08/2021	09/08/2021
					CT2727	Active implant devices/ancillaries	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem	09/08/2021	09/08/2021
									A070504	Premature Discharge of Battery	09/08/2021	09/08/2021
							F08	Hospitalization or Prolonged Hospitalization	A0705	Battery Problem	09/08/2021	09/08/2021
									A070504	Premature Discharge of Battery	09/08/2021	09/08/2021
							F1905	Device Revision or Replacement	A0705	Battery Problem	09/08/2021	09/08/2021
									A070504	Premature Discharge of Battery	09/08/2021	09/08/2021
					CT2750	Implantable cardiac electrical stimulators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem	09/08/2021	09/08/2021
									A070504	Premature Discharge of Battery	09/08/2021	09/08/2021
							F08	Hospitalization or Prolonged Hospitalization	A0705	Battery Problem	09/08/2021	09/08/2021
									A070504	Premature Discharge of Battery	09/08/2021	09/08/2021
							F1905	Device Revision or Replacement	A0705	Battery Problem	09/08/2021	09/08/2021
									A070504	Premature Discharge of Battery	09/08/2021	09/08/2021
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem	09/08/2021	09/08/2021
									A070504	Premature Discharge of Battery	09/08/2021	09/08/2021
							F08	Hospitalization or Prolonged Hospitalization	A0705	Battery Problem	09/08/2021	09/08/2021
									A070504	Premature Discharge of Battery	09/08/2021	09/08/2021
							F1905	Device Revision or Replacement	A0705	Battery Problem	09/08/2021	09/08/2021
									A070504	Premature Discharge of Battery	09/08/2021	09/08/2021
					CT2985	Active	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem	09/08/2021	09/08/2021
									A070504	Premature Discharge of Battery	09/08/2021	09/08/2021
							F08	Hospitalization or Prolonged Hospitalization	A0705	Battery Problem	09/08/2021	09/08/2021
									A070504	Premature Discharge of Battery	09/08/2021	09/08/2021
							F1905	Device Revision or Replacement	A0705	Battery Problem	09/08/2021	09/08/2021
									A070504	Premature Discharge of Battery	09/08/2021	09/08/2021
					CT3001	Implantable metal	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem	09/08/2021	09/08/2021
									A070504	Premature Discharge of Battery	09/08/2021	09/08/2021
							F08	Hospitalization or Prolonged Hospitalization	A0705	Battery Problem	09/08/2021	09/08/2021
									A070504	Premature Discharge of Battery	09/08/2021	09/08/2021
							F1905	Device Revision or Replacement	A0705	Battery Problem	09/08/2021	09/08/2021
									A070504	Premature Discharge of Battery	09/08/2021	09/08/2021
					CT3002	Implantable synthetic polymer	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem	09/08/2021	09/08/2021
									A070504	Premature Discharge of Battery	09/08/2021	09/08/2021
							F08	Hospitalization or Prolonged Hospitalization	A0705	Battery Problem	09/08/2021	09/08/2021
									A070504	Premature Discharge of Battery	09/08/2021	09/08/2021
							F1905	Device Revision or Replacement	A0705	Battery Problem	09/08/2021	09/08/2021
									A070504	Premature Discharge of Battery	09/08/2021	09/08/2021
					CT3006	Implantable inorganic material	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem	09/08/2021	09/08/2021
									A070504	Premature Discharge of Battery	09/08/2021	09/08/2021
							F08	Hospitalization or Prolonged Hospitalization	A0705	Battery Problem	09/08/2021	09/08/2021
									A070504	Premature Discharge of Battery	09/08/2021	09/08/2021
							F1905	Device Revision or Replacement	A0705	Battery Problem	09/08/2021	09/08/2021

									A070504	Premature Discharge of Battery	09/08/2021	09/08/2021
					CT314	Electrically powered	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem	09/08/2021	09/08/2021
									A070504	Premature Discharge of Battery	09/08/2021	09/08/2021
							F08	Hospitalization or Prolonged Hospitalization	A0705	Battery Problem	09/08/2021	09/08/2021
									A070504	Premature Discharge of Battery	09/08/2021	09/08/2021
							F1905	Device Revision or Replacement	A0705	Battery Problem	09/08/2021	09/08/2021
									A070504	Premature Discharge of Battery	09/08/2021	09/08/2021
					CT321	Long-term surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem	09/08/2021	09/08/2021
									A070504	Premature Discharge of Battery	09/08/2021	09/08/2021
							F08	Hospitalization or Prolonged Hospitalization	A0705	Battery Problem	09/08/2021	09/08/2021
									A070504	Premature Discharge of Battery	09/08/2021	09/08/2021
							F1905	Device Revision or Replacement	A0705	Battery Problem	09/08/2021	09/08/2021
									A070504	Premature Discharge of Battery	09/08/2021	09/08/2021
					CT449	Generators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem	09/08/2021	09/08/2021
									A070504	Premature Discharge of Battery	09/08/2021	09/08/2021
							F08	Hospitalization or Prolonged Hospitalization	A0705	Battery Problem	09/08/2021	09/08/2021
									A070504	Premature Discharge of Battery	09/08/2021	09/08/2021
							F1905	Device Revision or Replacement	A0705	Battery Problem	09/08/2021	09/08/2021
									A070504	Premature Discharge of Battery	09/08/2021	09/08/2021
					CT981	Single use	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem	09/08/2021	09/08/2021
									A070504	Premature Discharge of Battery	09/08/2021	09/08/2021
							F08	Hospitalization or Prolonged Hospitalization	A0705	Battery Problem	09/08/2021	09/08/2021
									A070504	Premature Discharge of Battery	09/08/2021	09/08/2021
							F1905	Device Revision or Replacement	A0705	Battery Problem	09/08/2021	09/08/2021
									A070504	Premature Discharge of Battery	09/08/2021	09/08/2021
					CT983	Surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem	09/08/2021	09/08/2021
									A070504	Premature Discharge of Battery	09/08/2021	09/08/2021
							F08	Hospitalization or Prolonged Hospitalization	A0705	Battery Problem	09/08/2021	09/08/2021
									A070504	Premature Discharge of Battery	09/08/2021	09/08/2021
							F1905	Device Revision or Replacement	A0705	Battery Problem	09/08/2021	09/08/2021
									A070504	Premature Discharge of Battery	09/08/2021	09/08/2021
6322	2021	All other reportable incidents	35852	Single-chamber implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem	06/02/2022	06/02/2022
									A070504	Premature Discharge of Battery	06/02/2022	06/02/2022
							F08	Hospitalization or Prolonged Hospitalization	A0705	Battery Problem	06/02/2022	06/02/2022
									A070504	Premature Discharge of Battery	06/02/2022	06/02/2022
							F1905	Device Revision or Replacement	A0705	Battery Problem	06/02/2022	06/02/2022
									A070504	Premature Discharge of Battery	06/02/2022	06/02/2022
6323	2021	All other reportable incidents	47265	Dual-chamber implantable pacemaker, rate-responsive	CT141	Electrodes/Leads and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	10/12/2021	10/12/2021
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	10/12/2021	10/12/2021
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	10/12/2021	10/12/2021
					CT145	Cardiovascular devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	10/12/2021	10/12/2021
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	10/12/2021	10/12/2021
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	10/12/2021	10/12/2021
					CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	10/12/2021	10/12/2021
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	10/12/2021	10/12/2021
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	10/12/2021	10/12/2021
					CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	10/12/2021	10/12/2021
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	10/12/2021	10/12/2021
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	10/12/2021	10/12/2021
					CT233	Surgical	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	10/12/2021	10/12/2021
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	10/12/2021	10/12/2021
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	10/12/2021	10/12/2021
					CT2406	Implantable	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	10/12/2021	10/12/2021
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	10/12/2021	10/12/2021
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	10/12/2021	10/12/2021
					CT2407	Active implantable	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	10/12/2021	10/12/2021
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	10/12/2021	10/12/2021
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	10/12/2021	10/12/2021
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	10/12/2021	10/12/2021
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	10/12/2021	10/12/2021
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	10/12/2021	10/12/2021
					CT2727	Active implant devices/ancillaries	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	10/12/2021	10/12/2021
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	10/12/2021	10/12/2021
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	10/12/2021	10/12/2021
					CT2750	Implantable cardiac electrical stimulators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	10/12/2021	10/12/2021
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	10/12/2021	10/12/2021
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	10/12/2021	10/12/2021
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	10/12/2021	10/12/2021
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	10/12/2021	10/12/2021
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	10/12/2021	10/12/2021
					CT2985	Active	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	10/12/2021	10/12/2021
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	10/12/2021	10/12/2021
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	10/12/2021	10/12/2021
					CT3001	Implantable metal	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	10/12/2021	10/12/2021
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	10/12/2021	10/12/2021
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	10/12/2021	10/12/2021
					CT3002	Implantable synthetic polymer	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	10/12/2021	10/12/2021
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	10/12/2021	10/12/2021
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	10/12/2021	10/12/2021
					CT3006	Implantable inorganic material	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	10/12/2021	10/12/2021
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	10/12/2021	10/12/2021
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	10/12/2021	10/12/2021
					CT314	Electrically powered	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	10/12/2021	10/12/2021
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	10/12/2021	10/12/2021
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	10/12/2021	10/12/2021
					CT321	Long-term surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	10/12/2021	10/12/2021
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	10/12/2021	10/12/2021
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	10/12/2021	10/12/2021

					CT449	Generators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	10/12/2021	10/12/2021
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	10/12/2021	10/12/2021
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	10/12/2021	10/12/2021
					CT981	Single use	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	10/12/2021	10/12/2021
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	10/12/2021	10/12/2021
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	10/12/2021	10/12/2021
					CT983	Surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	10/12/2021	10/12/2021
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	10/12/2021	10/12/2021
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	10/12/2021	10/12/2021
6324	2021	Death	47270	Cardiac resynchronization therapy implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A070909	Oversensing		
									A070910	Undersensing		
							F02	Death	A070909	Oversensing		
									A070910	Undersensing		
6325	2022	All other reportable incidents	35852	Single-chamber implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	25/03/2022	25/03/2022
							F1901	Additional Surgery	A070504	Premature Discharge of Battery	25/03/2022	25/03/2022
6326	2021	All other reportable incidents	47265	Dual-chamber implantable pacemaker, rate-responsive	CT141	Electrodes/Leads and associated devices	E011903	Syncope/Fainting	A05	Mechanical Problem		
									A070909	Oversensing		
									A0712	Pacing Problem		
									A072001	Electromagnetic Interference		
									A090801	Signal Artifact/Noise		
							E2403	No Clinical Signs, Symptoms or Conditions	A05	Mechanical Problem		
									A070909	Oversensing		
									A0712	Pacing Problem		
									A072001	Electromagnetic Interference		
									A090801	Signal Artifact/Noise		
							F08	Hospitalization or Prolonged Hospitalization	A05	Mechanical Problem		
									A070909	Oversensing		
									A0712	Pacing Problem		
									A072001	Electromagnetic Interference		
									A090801	Signal Artifact/Noise		
							F1203	Life Threatening Illness or Injury	A05	Mechanical Problem		
									A070909	Oversensing		
									A0712	Pacing Problem		
									A072001	Electromagnetic Interference		
									A090801	Signal Artifact/Noise		
							F1905	Device Revision or Replacement	A05	Mechanical Problem		
									A070909	Oversensing		
									A0712	Pacing Problem		
									A072001	Electromagnetic Interference		
									A090801	Signal Artifact/Noise		
					CT145	Cardiovascular devices	E011903	Syncope/Fainting	A05	Mechanical Problem		
									A070909	Oversensing		
									A0712	Pacing Problem		
									A072001	Electromagnetic Interference		
									A090801	Signal Artifact/Noise		
							E2403	No Clinical Signs, Symptoms or Conditions	A05	Mechanical Problem		
									A070909	Oversensing		
									A0712	Pacing Problem		
									A072001	Electromagnetic Interference		
									A090801	Signal Artifact/Noise		
							F08	Hospitalization or Prolonged Hospitalization	A05	Mechanical Problem		
									A070909	Oversensing		
									A0712	Pacing Problem		
									A072001	Electromagnetic Interference		
									A090801	Signal Artifact/Noise		
							F1203	Life Threatening Illness or Injury	A05	Mechanical Problem		
									A070909	Oversensing		
									A0712	Pacing Problem		
									A072001	Electromagnetic Interference		
									A090801	Signal Artifact/Noise		
							F1905	Device Revision or Replacement	A05	Mechanical Problem		
									A070909	Oversensing		
									A0712	Pacing Problem		
									A072001	Electromagnetic Interference		
									A090801	Signal Artifact/Noise		
					CT1684	Cardiac pacemakers and associated devices	E011903	Syncope/Fainting	A05	Mechanical Problem		
									A070909	Oversensing		
									A0712	Pacing Problem		
									A072001	Electromagnetic Interference		
									A090801	Signal Artifact/Noise		
							E2403	No Clinical Signs, Symptoms or Conditions	A05	Mechanical Problem		
									A070909	Oversensing		
									A0712	Pacing Problem		
									A072001	Electromagnetic Interference		
									A090801	Signal Artifact/Noise		
							F08	Hospitalization or Prolonged Hospitalization	A05	Mechanical Problem		
									A070909	Oversensing		
									A0712	Pacing Problem		
									A072001	Electromagnetic Interference		
									A090801	Signal Artifact/Noise		
							F1203	Life Threatening Illness or Injury	A05	Mechanical Problem		
									A070909	Oversensing		
									A0712	Pacing Problem		
									A072001	Electromagnetic Interference		
									A090801	Signal Artifact/Noise		
							F1905	Device Revision or Replacement	A05	Mechanical Problem		
									A070909	Oversensing		
									A0712	Pacing Problem		

									A072001	Electromagnetic Interference		
									A090801	Signal Artifact/Noise		
					CT1685	Cardiac pacemakers	E011903	Syncope/Fainting	A05	Mechanical Problem		
									A070909	Oversensing		
									A0712	Pacing Problem		
									A072001	Electromagnetic Interference		
									A090801	Signal Artifact/Noise		
							E2403	No Clinical Signs, Symptoms or Conditions	A05	Mechanical Problem		
									A070909	Oversensing		
									A0712	Pacing Problem		
									A072001	Electromagnetic Interference		
									A090801	Signal Artifact/Noise		
							F08	Hospitalization or Prolonged Hospitalization	A05	Mechanical Problem		
									A070909	Oversensing		
									A0712	Pacing Problem		
									A072001	Electromagnetic Interference		
									A090801	Signal Artifact/Noise		
							F1203	Life Threatening Illness or Injury	A05	Mechanical Problem		
									A070909	Oversensing		
									A0712	Pacing Problem		
									A072001	Electromagnetic Interference		
									A090801	Signal Artifact/Noise		
							F1905	Device Revision or Replacement	A05	Mechanical Problem		
									A070909	Oversensing		
									A0712	Pacing Problem		
									A072001	Electromagnetic Interference		
									A090801	Signal Artifact/Noise		
					CT233	Surgical	E011903	Syncope/Fainting	A05	Mechanical Problem		
									A070909	Oversensing		
									A0712	Pacing Problem		
									A072001	Electromagnetic Interference		
									A090801	Signal Artifact/Noise		
							E2403	No Clinical Signs, Symptoms or Conditions	A05	Mechanical Problem		
									A070909	Oversensing		
									A0712	Pacing Problem		
									A072001	Electromagnetic Interference		
									A090801	Signal Artifact/Noise		
							F08	Hospitalization or Prolonged Hospitalization	A05	Mechanical Problem		
									A070909	Oversensing		
									A0712	Pacing Problem		
									A072001	Electromagnetic Interference		
									A090801	Signal Artifact/Noise		
							F1203	Life Threatening Illness or Injury	A05	Mechanical Problem		
									A070909	Oversensing		
									A0712	Pacing Problem		
									A072001	Electromagnetic Interference		
									A090801	Signal Artifact/Noise		
							F1905	Device Revision or Replacement	A05	Mechanical Problem		
									A070909	Oversensing		
									A0712	Pacing Problem		
									A072001	Electromagnetic Interference		
									A090801	Signal Artifact/Noise		
					CT2406	Implantable	E011903	Syncope/Fainting	A05	Mechanical Problem		
									A070909	Oversensing		
									A0712	Pacing Problem		
									A072001	Electromagnetic Interference		
									A090801	Signal Artifact/Noise		
							E2403	No Clinical Signs, Symptoms or Conditions	A05	Mechanical Problem		
									A070909	Oversensing		
									A0712	Pacing Problem		
									A072001	Electromagnetic Interference		
									A090801	Signal Artifact/Noise		
							F08	Hospitalization or Prolonged Hospitalization	A05	Mechanical Problem		
									A070909	Oversensing		
									A0712	Pacing Problem		
									A072001	Electromagnetic Interference		
									A090801	Signal Artifact/Noise		
							F1203	Life Threatening Illness or Injury	A05	Mechanical Problem		
									A070909	Oversensing		
									A0712	Pacing Problem		
									A072001	Electromagnetic Interference		
									A090801	Signal Artifact/Noise		
							F1905	Device Revision or Replacement	A05	Mechanical Problem		
									A070909	Oversensing		
									A0712	Pacing Problem		
									A072001	Electromagnetic Interference		
									A090801	Signal Artifact/Noise		
					CT2407	Active implantable	E011903	Syncope/Fainting	A05	Mechanical Problem		
									A070909	Oversensing		
									A0712	Pacing Problem		
									A072001	Electromagnetic Interference		
									A090801	Signal Artifact/Noise		
							E2403	No Clinical Signs, Symptoms or Conditions	A05	Mechanical Problem		
									A070909	Oversensing		
									A0712	Pacing Problem		
									A072001	Electromagnetic Interference		
									A090801	Signal Artifact/Noise		
							F08	Hospitalization or Prolonged Hospitalization	A05	Mechanical Problem		

									A070909	Oversensing			
									A0712	Pacing Problem			
									A072001	Electromagnetic Interference			
									A090801	Signal Artifact/Noise			
								F1203	Life Threatening Illness or Injury	A05	Mechanical Problem		
										A070909	Oversensing		
										A0712	Pacing Problem		
										A072001	Electromagnetic Interference		
										A090801	Signal Artifact/Noise		
								F1905	Device Revision or Replacement	A05	Mechanical Problem		
										A070909	Oversensing		
										A0712	Pacing Problem		
										A072001	Electromagnetic Interference		
										A090801	Signal Artifact/Noise		
							CT257	Implantable cardiac pacemakers	E011903	Syncope/Fainting	A05	Mechanical Problem	
											A070909	Oversensing	
											A0712	Pacing Problem	
											A072001	Electromagnetic Interference	
											A090801	Signal Artifact/Noise	
									E2403	No Clinical Signs, Symptoms or Conditions	A05	Mechanical Problem	
											A070909	Oversensing	
											A0712	Pacing Problem	
											A072001	Electromagnetic Interference	
											A090801	Signal Artifact/Noise	
									F08	Hospitalization or Prolonged Hospitalization	A05	Mechanical Problem	
											A070909	Oversensing	
											A0712	Pacing Problem	
											A072001	Electromagnetic Interference	
											A090801	Signal Artifact/Noise	
									F1203	Life Threatening Illness or Injury	A05	Mechanical Problem	
											A070909	Oversensing	
											A0712	Pacing Problem	
											A072001	Electromagnetic Interference	
											A090801	Signal Artifact/Noise	
									F1905	Device Revision or Replacement	A05	Mechanical Problem	
											A070909	Oversensing	
											A0712	Pacing Problem	
											A072001	Electromagnetic Interference	
											A090801	Signal Artifact/Noise	
							CT2727	Active implant devices/ancillaries	E011903	Syncope/Fainting	A05	Mechanical Problem	
											A070909	Oversensing	
											A0712	Pacing Problem	
											A072001	Electromagnetic Interference	
											A090801	Signal Artifact/Noise	
									E2403	No Clinical Signs, Symptoms or Conditions	A05	Mechanical Problem	
											A070909	Oversensing	
											A0712	Pacing Problem	
											A072001	Electromagnetic Interference	
											A090801	Signal Artifact/Noise	
									F08	Hospitalization or Prolonged Hospitalization	A05	Mechanical Problem	
											A070909	Oversensing	
											A0712	Pacing Problem	
											A072001	Electromagnetic Interference	
											A090801	Signal Artifact/Noise	
									F1203	Life Threatening Illness or Injury	A05	Mechanical Problem	
											A070909	Oversensing	
											A0712	Pacing Problem	
											A072001	Electromagnetic Interference	
											A090801	Signal Artifact/Noise	
									F1905	Device Revision or Replacement	A05	Mechanical Problem	
											A070909	Oversensing	
											A0712	Pacing Problem	
											A072001	Electromagnetic Interference	
											A090801	Signal Artifact/Noise	
							CT2750	Implantable cardiac electrical stimulators and associated devices	E011903	Syncope/Fainting	A05	Mechanical Problem	
											A070909	Oversensing	
											A0712	Pacing Problem	
											A072001	Electromagnetic Interference	
											A090801	Signal Artifact/Noise	
									E2403	No Clinical Signs, Symptoms or Conditions	A05	Mechanical Problem	
											A070909	Oversensing	
											A0712	Pacing Problem	
											A072001	Electromagnetic Interference	
											A090801	Signal Artifact/Noise	
									F08	Hospitalization or Prolonged Hospitalization	A05	Mechanical Problem	
											A070909	Oversensing	
											A0712	Pacing Problem	
											A072001	Electromagnetic Interference	
											A090801	Signal Artifact/Noise	
									F1203	Life Threatening Illness or Injury	A05	Mechanical Problem	
											A070909	Oversensing	
											A0712	Pacing Problem	
											A072001	Electromagnetic Interference	
											A090801	Signal Artifact/Noise	
									F1905	Device Revision or Replacement	A05	Mechanical Problem	
											A070909	Oversensing	
											A0712	Pacing Problem	
											A072001	Electromagnetic Interference	

								A090801	Signal Artifact/Noise		
				CT2866	Implantable cardiac electrical stimulators	E011903	Syncope/Fainting	A05	Mechanical Problem		
								A070909	Oversensing		
								A0712	Pacing Problem		
								A072001	Electromagnetic Interference		
								A090801	Signal Artifact/Noise		
						E2403	No Clinical Signs, Symptoms or Conditions	A05	Mechanical Problem		
								A070909	Oversensing		
								A0712	Pacing Problem		
								A072001	Electromagnetic Interference		
								A090801	Signal Artifact/Noise		
						F08	Hospitalization or Prolonged Hospitalization	A05	Mechanical Problem		
								A070909	Oversensing		
								A0712	Pacing Problem		
								A072001	Electromagnetic Interference		
								A090801	Signal Artifact/Noise		
						F1203	Life Threatening Illness or Injury	A05	Mechanical Problem		
								A070909	Oversensing		
								A0712	Pacing Problem		
								A072001	Electromagnetic Interference		
								A090801	Signal Artifact/Noise		
						F1905	Device Revision or Replacement	A05	Mechanical Problem		
								A070909	Oversensing		
								A0712	Pacing Problem		
								A072001	Electromagnetic Interference		
								A090801	Signal Artifact/Noise		
				CT2985	Active	E011903	Syncope/Fainting	A05	Mechanical Problem		
								A070909	Oversensing		
								A0712	Pacing Problem		
								A072001	Electromagnetic Interference		
								A090801	Signal Artifact/Noise		
						E2403	No Clinical Signs, Symptoms or Conditions	A05	Mechanical Problem		
								A070909	Oversensing		
								A0712	Pacing Problem		
								A072001	Electromagnetic Interference		
								A090801	Signal Artifact/Noise		
						F08	Hospitalization or Prolonged Hospitalization	A05	Mechanical Problem		
								A070909	Oversensing		
								A0712	Pacing Problem		
								A072001	Electromagnetic Interference		
								A090801	Signal Artifact/Noise		
						F1203	Life Threatening Illness or Injury	A05	Mechanical Problem		
								A070909	Oversensing		
								A0712	Pacing Problem		
								A072001	Electromagnetic Interference		
								A090801	Signal Artifact/Noise		
						F1905	Device Revision or Replacement	A05	Mechanical Problem		
								A070909	Oversensing		
								A0712	Pacing Problem		
								A072001	Electromagnetic Interference		
								A090801	Signal Artifact/Noise		
				CT3001	Implantable metal	E011903	Syncope/Fainting	A05	Mechanical Problem		
								A070909	Oversensing		
								A0712	Pacing Problem		
								A072001	Electromagnetic Interference		
								A090801	Signal Artifact/Noise		
						E2403	No Clinical Signs, Symptoms or Conditions	A05	Mechanical Problem		
								A070909	Oversensing		
								A0712	Pacing Problem		
								A072001	Electromagnetic Interference		
								A090801	Signal Artifact/Noise		
						F08	Hospitalization or Prolonged Hospitalization	A05	Mechanical Problem		
								A070909	Oversensing		
								A0712	Pacing Problem		
								A072001	Electromagnetic Interference		
								A090801	Signal Artifact/Noise		
						F1203	Life Threatening Illness or Injury	A05	Mechanical Problem		
								A070909	Oversensing		
								A0712	Pacing Problem		
								A072001	Electromagnetic Interference		
								A090801	Signal Artifact/Noise		
						F1905	Device Revision or Replacement	A05	Mechanical Problem		
								A070909	Oversensing		
								A0712	Pacing Problem		
								A072001	Electromagnetic Interference		
								A090801	Signal Artifact/Noise		
				CT3002	Implantable synthetic polymer	E011903	Syncope/Fainting	A05	Mechanical Problem		
								A070909	Oversensing		
								A0712	Pacing Problem		
								A072001	Electromagnetic Interference		
								A090801	Signal Artifact/Noise		
						E2403	No Clinical Signs, Symptoms or Conditions	A05	Mechanical Problem		
								A070909	Oversensing		
								A0712	Pacing Problem		
								A072001	Electromagnetic Interference		
								A090801	Signal Artifact/Noise		
						F08	Hospitalization or Prolonged Hospitalization	A05	Mechanical Problem		
								A070909	Oversensing		

									A0712	Pacing Problem		
									A072001	Electromagnetic Interference		
									A090801	Signal Artifact/Noise		
								F1203	Life Threatening Illness or Injury	A05	Mechanical Problem		
										A070909	Oversensing		
										A0712	Pacing Problem		
										A072001	Electromagnetic Interference		
										A090801	Signal Artifact/Noise		
								F1905	Device Revision or Replacement	A05	Mechanical Problem		
										A070909	Oversensing		
										A0712	Pacing Problem		
										A072001	Electromagnetic Interference		
										A090801	Signal Artifact/Noise		
							CT3006	Implantable inorganic material	E011903	Syncope/Fainting	A05	Mechanical Problem		
											A070909	Oversensing		
											A0712	Pacing Problem		
											A072001	Electromagnetic Interference		
											A090801	Signal Artifact/Noise		
									E2403	No Clinical Signs, Symptoms or Conditions	A05	Mechanical Problem		
											A070909	Oversensing		
											A0712	Pacing Problem		
											A072001	Electromagnetic Interference		
											A090801	Signal Artifact/Noise		
									F08	Hospitalization or Prolonged Hospitalization	A05	Mechanical Problem		
											A070909	Oversensing		
											A0712	Pacing Problem		
											A072001	Electromagnetic Interference		
											A090801	Signal Artifact/Noise		
									F1203	Life Threatening Illness or Injury	A05	Mechanical Problem		
											A070909	Oversensing		
											A0712	Pacing Problem		
											A072001	Electromagnetic Interference		
											A090801	Signal Artifact/Noise		
									F1905	Device Revision or Replacement	A05	Mechanical Problem		
											A070909	Oversensing		
											A0712	Pacing Problem		
											A072001	Electromagnetic Interference		
											A090801	Signal Artifact/Noise		
								CT314	Electrically powered	E011903	Syncope/Fainting	A05	Mechanical Problem		
												A070909	Oversensing		
												A0712	Pacing Problem		
												A072001	Electromagnetic Interference		
												A090801	Signal Artifact/Noise		
										E2403	No Clinical Signs, Symptoms or Conditions	A05	Mechanical Problem		
												A070909	Oversensing		
												A0712	Pacing Problem		
												A072001	Electromagnetic Interference		
												A090801	Signal Artifact/Noise		
										F08	Hospitalization or Prolonged Hospitalization	A05	Mechanical Problem		
												A070909	Oversensing		
												A0712	Pacing Problem		
												A072001	Electromagnetic Interference		
												A090801	Signal Artifact/Noise		
										F1203	Life Threatening Illness or Injury	A05	Mechanical Problem		
												A070909	Oversensing		
												A0712	Pacing Problem		
												A072001	Electromagnetic Interference		
												A090801	Signal Artifact/Noise		
										F1905	Device Revision or Replacement	A05	Mechanical Problem		
												A070909	Oversensing		
												A0712	Pacing Problem		
												A072001	Electromagnetic Interference		
												A090801	Signal Artifact/Noise		
								CT321	Long-term surgical invasive	E011903	Syncope/Fainting	A05	Mechanical Problem		
												A070909	Oversensing		
												A0712	Pacing Problem		
												A072001	Electromagnetic Interference		
												A090801	Signal Artifact/Noise		
										E2403	No Clinical Signs, Symptoms or Conditions	A05	Mechanical Problem		
												A070909	Oversensing		
												A0712	Pacing Problem		
												A072001	Electromagnetic Interference		
												A090801	Signal Artifact/Noise		
										F08	Hospitalization or Prolonged Hospitalization	A05	Mechanical Problem		
												A070909	Oversensing		
												A0712	Pacing Problem		
												A072001	Electromagnetic Interference		
												A090801	Signal Artifact/Noise		
										F1203	Life Threatening Illness or Injury	A05	Mechanical Problem		
												A070909	Oversensing		
												A0712	Pacing Problem		
												A072001	Electromagnetic Interference		
												A090801	Signal Artifact/Noise		
										F1905	Device Revision or Replacement	A05	Mechanical Problem		
												A070909	Oversensing		
												A0712	Pacing Problem		
												A072001	Electromagnetic Interference		
												A090801	Signal Artifact/Noise		

					CT449	Generators and associated devices	E011903	Syncope/Fainting	A05	Mechanical Problem		
									A070909	Oversensing		
									A0712	Pacing Problem		
									A072001	Electromagnetic Interference		
									A090801	Signal Artifact/Noise		
							E2403	No Clinical Signs, Symptoms or Conditions	A05	Mechanical Problem		
									A070909	Oversensing		
									A0712	Pacing Problem		
									A072001	Electromagnetic Interference		
									A090801	Signal Artifact/Noise		
							F08	Hospitalization or Prolonged Hospitalization	A05	Mechanical Problem		
									A070909	Oversensing		
									A0712	Pacing Problem		
									A072001	Electromagnetic Interference		
									A090801	Signal Artifact/Noise		
							F1203	Life Threatening Illness or Injury	A05	Mechanical Problem		
									A070909	Oversensing		
									A0712	Pacing Problem		
									A072001	Electromagnetic Interference		
									A090801	Signal Artifact/Noise		
							F1905	Device Revision or Replacement	A05	Mechanical Problem		
									A070909	Oversensing		
									A0712	Pacing Problem		
									A072001	Electromagnetic Interference		
									A090801	Signal Artifact/Noise		
					CT981	Single use	E011903	Syncope/Fainting	A05	Mechanical Problem		
									A070909	Oversensing		
									A0712	Pacing Problem		
									A072001	Electromagnetic Interference		
									A090801	Signal Artifact/Noise		
							E2403	No Clinical Signs, Symptoms or Conditions	A05	Mechanical Problem		
									A070909	Oversensing		
									A0712	Pacing Problem		
									A072001	Electromagnetic Interference		
									A090801	Signal Artifact/Noise		
							F08	Hospitalization or Prolonged Hospitalization	A05	Mechanical Problem		
									A070909	Oversensing		
									A0712	Pacing Problem		
									A072001	Electromagnetic Interference		
									A090801	Signal Artifact/Noise		
							F1203	Life Threatening Illness or Injury	A05	Mechanical Problem		
									A070909	Oversensing		
									A0712	Pacing Problem		
									A072001	Electromagnetic Interference		
									A090801	Signal Artifact/Noise		
							F1905	Device Revision or Replacement	A05	Mechanical Problem		
									A070909	Oversensing		
									A0712	Pacing Problem		
									A072001	Electromagnetic Interference		
									A090801	Signal Artifact/Noise		
					CT983	Surgical invasive	E011903	Syncope/Fainting	A05	Mechanical Problem		
									A070909	Oversensing		
									A0712	Pacing Problem		
									A072001	Electromagnetic Interference		
									A090801	Signal Artifact/Noise		
							E2403	No Clinical Signs, Symptoms or Conditions	A05	Mechanical Problem		
									A070909	Oversensing		
									A0712	Pacing Problem		
									A072001	Electromagnetic Interference		
									A090801	Signal Artifact/Noise		
							F08	Hospitalization or Prolonged Hospitalization	A05	Mechanical Problem		
									A070909	Oversensing		
									A0712	Pacing Problem		
									A072001	Electromagnetic Interference		
									A090801	Signal Artifact/Noise		
							F1203	Life Threatening Illness or Injury	A05	Mechanical Problem		
									A070909	Oversensing		
									A0712	Pacing Problem		
									A072001	Electromagnetic Interference		
									A090801	Signal Artifact/Noise		
							F1905	Device Revision or Replacement	A05	Mechanical Problem		
									A070909	Oversensing		
									A0712	Pacing Problem		
									A072001	Electromagnetic Interference		
									A090801	Signal Artifact/Noise		
6327	2022	All other reportable incidents	47265	Dual-chamber implantable pacemaker, rate-responsive	CT141	Electrodes/Leads and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	11/07/2022	11/07/2022
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	11/07/2022	11/07/2022
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	11/07/2022	11/07/2022
					CT145	Cardiovascular devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	11/07/2022	11/07/2022
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	11/07/2022	11/07/2022
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	11/07/2022	11/07/2022
					CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	11/07/2022	11/07/2022
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	11/07/2022	11/07/2022
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	11/07/2022	11/07/2022
					CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	11/07/2022	11/07/2022
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	11/07/2022	11/07/2022
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	11/07/2022	11/07/2022
					CT233	Surgical	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	11/07/2022	11/07/2022

							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	11/07/2022	11/07/2022
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	11/07/2022	11/07/2022
					CT2406	Implantable	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	11/07/2022	11/07/2022
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	11/07/2022	11/07/2022
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	11/07/2022	11/07/2022
					CT2407	Active implantable	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	11/07/2022	11/07/2022
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	11/07/2022	11/07/2022
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	11/07/2022	11/07/2022
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	11/07/2022	11/07/2022
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	11/07/2022	11/07/2022
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	11/07/2022	11/07/2022
					CT2727	Active implant devices/ancillaries	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	11/07/2022	11/07/2022
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	11/07/2022	11/07/2022
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	11/07/2022	11/07/2022
					CT2750	Implantable cardiac electrical stimulators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	11/07/2022	11/07/2022
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	11/07/2022	11/07/2022
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	11/07/2022	11/07/2022
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	11/07/2022	11/07/2022
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	11/07/2022	11/07/2022
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	11/07/2022	11/07/2022
					CT2985	Active	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	11/07/2022	11/07/2022
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	11/07/2022	11/07/2022
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	11/07/2022	11/07/2022
					CT3001	Implantable metal	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	11/07/2022	11/07/2022
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	11/07/2022	11/07/2022
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	11/07/2022	11/07/2022
					CT3002	Implantable synthetic polymer	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	11/07/2022	11/07/2022
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	11/07/2022	11/07/2022
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	11/07/2022	11/07/2022
					CT3006	Implantable inorganic material	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	11/07/2022	11/07/2022
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	11/07/2022	11/07/2022
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	11/07/2022	11/07/2022
					CT314	Electrically powered	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	11/07/2022	11/07/2022
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	11/07/2022	11/07/2022
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	11/07/2022	11/07/2022
					CT321	Long-term surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	11/07/2022	11/07/2022
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	11/07/2022	11/07/2022
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	11/07/2022	11/07/2022
					CT449	Generators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	11/07/2022	11/07/2022
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	11/07/2022	11/07/2022
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	11/07/2022	11/07/2022
					CT981	Single use	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	11/07/2022	11/07/2022
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	11/07/2022	11/07/2022
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	11/07/2022	11/07/2022
					CT983	Surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	11/07/2022	11/07/2022
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	11/07/2022	11/07/2022
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	11/07/2022	11/07/2022
6328	2022	All other reportable incidents	47265	Dual-chamber implantable pacemaker, rate-responsive	CT141	Electrodes/Leads and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
					CT145	Cardiovascular devices	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
					CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
					CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
					CT233	Surgical	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
					CT2406	Implantable	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
					CT2407	Active implantable	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
					CT2727	Active implant devices/ancillaries	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
					CT2750	Implantable cardiac electrical stimulators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem		

									A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
					CT2985	Active	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
					CT3001	Implantable metal	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
					CT3002	Implantable synthetic polymer	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
					CT3006	Implantable inorganic material	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
					CT314	Electrically powered	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
					CT321	Long-term surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
					CT449	Generators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
					CT981	Single use	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
					CT983	Surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
6329	2022	All other reportable incidents	35852	Single-chamber implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
6330	2022	All other reportable incidents	37265	Dual-chamber implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E2311	Discomfort	A0908	Incorrect, Inadequate or Imprecise Result or Readings	30/03/2022	30/03/2022
							E2403	No Clinical Signs, Symptoms or Conditions	A0908	Incorrect, Inadequate or Imprecise Result or Readings	30/03/2022	30/03/2022
							F08	Hospitalization or Prolonged Hospitalization	A0908	Incorrect, Inadequate or Imprecise Result or Readings	30/03/2022	30/03/2022
							F1903	Device Explantation	A0908	Incorrect, Inadequate or Imprecise Result or Readings	30/03/2022	30/03/2022
6331	2022	All other reportable incidents	36241	Pacemaker/defibrillator lead	CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070101	Failure to Capture		
									A072201	High Impedance		
							F08	Hospitalization or Prolonged Hospitalization	A070101	Failure to Capture		
									A072201	High Impedance		
							F1001	Absence of Treatment	A070101	Failure to Capture		
									A072201	High Impedance		
							F1905	Device Revision or Replacement	A070101	Failure to Capture		
									A072201	High Impedance		
							F1909	Surgical Procedure Delayed	A070101	Failure to Capture		
									A072201	High Impedance		
6332	2021	All other reportable incidents	35852	Single-chamber implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E2104	Electric Shock	A0709	Device Sensing Problem		
									A070909	Oversensing		
									A071302	Inappropriate/Inadequate Shock/Stimulation		
							E2403	No Clinical Signs, Symptoms or Conditions	A0709	Device Sensing Problem		
									A070909	Oversensing		
									A071302	Inappropriate/Inadequate Shock/Stimulation		
							F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A0709	Device Sensing Problem		
									A070909	Oversensing		
									A071302	Inappropriate/Inadequate Shock/Stimulation		
6333	2021	All other reportable incidents	47265	Dual-chamber implantable pacemaker, rate-responsive	CT141	Electrodes/Leads and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	29/10/2021	29/10/2021
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	29/10/2021	29/10/2021
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	29/10/2021	29/10/2021
					CT145	Cardiovascular devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	29/10/2021	29/10/2021
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	29/10/2021	29/10/2021
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	29/10/2021	29/10/2021
					CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	29/10/2021	29/10/2021
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	29/10/2021	29/10/2021
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	29/10/2021	29/10/2021
					CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	29/10/2021	29/10/2021
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	29/10/2021	29/10/2021
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	29/10/2021	29/10/2021
					CT233	Surgical	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	29/10/2021	29/10/2021
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	29/10/2021	29/10/2021
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	29/10/2021	29/10/2021
					CT2406	Implantable	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	29/10/2021	29/10/2021
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	29/10/2021	29/10/2021
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	29/10/2021	29/10/2021
					CT2407	Active implantable	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	29/10/2021	29/10/2021

							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	29/10/2021	29/10/2021
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	29/10/2021	29/10/2021
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	29/10/2021	29/10/2021
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	29/10/2021	29/10/2021
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	29/10/2021	29/10/2021
					CT2727	Active implant devices/ancillaries	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	29/10/2021	29/10/2021
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	29/10/2021	29/10/2021
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	29/10/2021	29/10/2021
					CT2750	Implantable cardiac electrical stimulators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	29/10/2021	29/10/2021
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	29/10/2021	29/10/2021
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	29/10/2021	29/10/2021
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	29/10/2021	29/10/2021
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	29/10/2021	29/10/2021
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	29/10/2021	29/10/2021
					CT2985	Active	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	29/10/2021	29/10/2021
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	29/10/2021	29/10/2021
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	29/10/2021	29/10/2021
					CT3001	Implantable metal	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	29/10/2021	29/10/2021
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	29/10/2021	29/10/2021
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	29/10/2021	29/10/2021
					CT3002	Implantable synthetic polymer	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	29/10/2021	29/10/2021
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	29/10/2021	29/10/2021
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	29/10/2021	29/10/2021
					CT3006	Implantable inorganic material	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	29/10/2021	29/10/2021
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	29/10/2021	29/10/2021
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	29/10/2021	29/10/2021
					CT314	Electrically powered	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	29/10/2021	29/10/2021
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	29/10/2021	29/10/2021
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	29/10/2021	29/10/2021
					CT321	Long-term surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	29/10/2021	29/10/2021
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	29/10/2021	29/10/2021
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	29/10/2021	29/10/2021
					CT449	Generators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	29/10/2021	29/10/2021
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	29/10/2021	29/10/2021
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	29/10/2021	29/10/2021
					CT981	Single use	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	29/10/2021	29/10/2021
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	29/10/2021	29/10/2021
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	29/10/2021	29/10/2021
					CT983	Surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	29/10/2021	29/10/2021
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	29/10/2021	29/10/2021
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	29/10/2021	29/10/2021
6334	2021	All other reportable incidents	36241	Pacemaker/defibrillator lead	CT1684	Cardiac pacemakers and associated devices	E0605	Cardiac Tamponade	A051201	Device Dislodged or Dislocated		
									A0709	Device Sensing Problem		
							E0619	Pericardial Effusion	A051201	Device Dislodged or Dislocated		
									A0709	Device Sensing Problem		
							E0717	Dyspnea	A051201	Device Dislodged or Dislocated		
									A0709	Device Sensing Problem		
							E2114	Perforation	A051201	Device Dislodged or Dislocated		
									A0709	Device Sensing Problem		
							E2403	No Clinical Signs, Symptoms or Conditions	A051201	Device Dislodged or Dislocated		
									A0709	Device Sensing Problem		
							F08	Hospitalization or Prolonged Hospitalization	A051201	Device Dislodged or Dislocated		
									A0709	Device Sensing Problem		
							F1905	Device Revision or Replacement	A051201	Device Dislodged or Dislocated		
									A0709	Device Sensing Problem		
6335	2021	All other reportable incidents	35852	Single-chamber implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E2104	Electric Shock	A071302	Inappropriate/Inadequate Shock/Stimulation		
							E2403	No Clinical Signs, Symptoms or Conditions	A071302	Inappropriate/Inadequate Shock/Stimulation		
							F08	Hospitalization or Prolonged Hospitalization	A071302	Inappropriate/Inadequate Shock/Stimulation		
6336	2021	All other reportable incidents	35852	Single-chamber implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem	13/08/2021	13/08/2021
									A070504	Premature Discharge of Battery	13/08/2021	13/08/2021
							F08	Hospitalization or Prolonged Hospitalization	A0705	Battery Problem	13/08/2021	13/08/2021
									A070504	Premature Discharge of Battery	13/08/2021	13/08/2021
							F1905	Device Revision or Replacement	A0705	Battery Problem	13/08/2021	13/08/2021
									A070504	Premature Discharge of Battery	13/08/2021	13/08/2021
6337	2021	All other reportable incidents	36241	Pacemaker/defibrillator lead	CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A072201	High Impedance		
							F26	No Health Consequences or Impact	A072201	High Impedance		
6338	2021	All other reportable incidents	47265	Dual-chamber implantable pacemaker, rate-responsive	CT141	Electrodes/Leads and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT145	Cardiovascular devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT233	Surgical	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT2406	Implantable	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT2407	Active implantable	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT2727	Active implant devices/ancillaries	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT2750	Implantable cardiac electrical stimulators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		

					CT2985	Active	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT3001	Implantable metal	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT3002	Implantable synthetic polymer	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT3006	Implantable inorganic material	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT314	Electrically powered	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT321	Long-term surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT449	Generators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT981	Single use	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT983	Surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
6339	2022	All other reportable incidents	47265	Dual-chamber implantable pacemaker, rate-responsive	CT141	Electrodes/Leads and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT145	Cardiovascular devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT233	Surgical	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT2406	Implantable	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT2407	Active implantable	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT2727	Active implant devices/ancillaries	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT2750	Implantable cardiac electrical stimulators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT2985	Active	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT3001	Implantable metal	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT3002	Implantable synthetic polymer	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT3006	Implantable inorganic material	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT314	Electrically powered	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT321	Long-term surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT449	Generators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT981	Single use	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT983	Surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
6340	2021	All other reportable incidents	37265	Dual-chamber implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A070909	Oversensing		
									A0712	Pacing Problem		
							F26	No Health Consequences or Impact	A070909	Oversensing		
									A0712	Pacing Problem		
6341	2021	Death	35852	Single-chamber implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E0602	Cardiac Arrest	A070909	Oversensing	06/08/2021	06/08/2021
									A071201	Failure to Convert Rhythm	06/08/2021	06/08/2021
									A071302	Inappropriate/Inadequate Shock/Stimulation	06/08/2021	06/08/2021
									A072201	High Impedance	06/08/2021	06/08/2021
							F02	Death	A070909	Oversensing	06/08/2021	06/08/2021
									A071201	Failure to Convert Rhythm	06/08/2021	06/08/2021
									A071302	Inappropriate/Inadequate Shock/Stimulation	06/08/2021	06/08/2021
									A072201	High Impedance	06/08/2021	06/08/2021
							F2306	Resuscitation	A070909	Oversensing	06/08/2021	06/08/2021
									A071201	Failure to Convert Rhythm	06/08/2021	06/08/2021
									A071302	Inappropriate/Inadequate Shock/Stimulation	06/08/2021	06/08/2021
									A072201	High Impedance	06/08/2021	06/08/2021
6342	2021	Death	36241	Pacemaker/defibrillator lead	CT1684	Cardiac pacemakers and associated devices	E0602	Cardiac Arrest	A070909	Oversensing		
									A071201	Failure to Convert Rhythm		
									A071302	Inappropriate/Inadequate Shock/Stimulation		
									A072201	High Impedance		
							F02	Death	A070909	Oversensing		
									A071201	Failure to Convert Rhythm		
									A071302	Inappropriate/Inadequate Shock/Stimulation		
									A072201	High Impedance		
							F1904	Device Repositioning	A070909	Oversensing		
									A071201	Failure to Convert Rhythm		
									A071302	Inappropriate/Inadequate Shock/Stimulation		
									A072201	High Impedance		
							F2306	Resuscitation	A070909	Oversensing		
									A071201	Failure to Convert Rhythm		

									A071302	Inappropriate/Inadequate Shock/Stimulation		
									A072201	High Impedance		
6343	2021	All other reportable incidents	63188	Left-ventricular wireless pacing system sensor/transmitter	CT141	Electrodes/Leads and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem		
					CT145	Cardiovascular devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem		
					CT1577	Wireless monitoring system transmitters	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem		
					CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem		
					CT2317	Transmitters/Receivers/Transceivers	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem		
					CT233	Surgical	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem		
					CT2406	Implantable	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem		
					CT2407	Active implantable	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem		
					CT255	Transducers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem		
					CT2726	Physiological monitoring/measuring devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem		
					CT2727	Active implant devices/ancillaries	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem		
					CT2750	Implantable cardiac electrical stimulators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem		
					CT2885	Physiological monitoring/measuring system transmitters/receivers/transceivers	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem		
					CT2985	Active	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem		
					CT3001	Implantable metal	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem		
					CT3002	Implantable synthetic polymer	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem		
					CT3006	Implantable inorganic material	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem		
					CT314	Electrically powered	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem		
					CT321	Long-term surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem		
					CT384	Sensors and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem		
					CT449	Generators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem		
					CT981	Single use	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem		
					CT983	Surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem		
					CT994	Sensors	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem		
6344	2022	All other reportable incidents	47265	Dual-chamber implantable pacemaker, rate-responsive	CT141	Electrodes/Leads and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070101	Failure to Capture		
							F26	No Health Consequences or Impact	A070101	Failure to Capture		
					CT145	Cardiovascular devices	E2403	No Clinical Signs, Symptoms or Conditions	A070101	Failure to Capture		
							F26	No Health Consequences or Impact	A070101	Failure to Capture		
					CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070101	Failure to Capture		
							F26	No Health Consequences or Impact	A070101	Failure to Capture		
					CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A070101	Failure to Capture		
							F26	No Health Consequences or Impact	A070101	Failure to Capture		
					CT233	Surgical	E2403	No Clinical Signs, Symptoms or Conditions	A070101	Failure to Capture		
							F26	No Health Consequences or Impact	A070101	Failure to Capture		
					CT2406	Implantable	E2403	No Clinical Signs, Symptoms or Conditions	A070101	Failure to Capture		
							F26	No Health Consequences or Impact	A070101	Failure to Capture		
					CT2407	Active implantable	E2403	No Clinical Signs, Symptoms or Conditions	A070101	Failure to Capture		
							F26	No Health Consequences or Impact	A070101	Failure to Capture		
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A070101	Failure to Capture		
							F26	No Health Consequences or Impact	A070101	Failure to Capture		
					CT2727	Active implant devices/ancillaries	E2403	No Clinical Signs, Symptoms or Conditions	A070101	Failure to Capture		
							F26	No Health Consequences or Impact	A070101	Failure to Capture		
					CT2750	Implantable cardiac electrical stimulators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070101	Failure to Capture		
							F26	No Health Consequences or Impact	A070101	Failure to Capture		
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A070101	Failure to Capture		
							F26	No Health Consequences or Impact	A070101	Failure to Capture		
					CT2985	Active	E2403	No Clinical Signs, Symptoms or Conditions	A070101	Failure to Capture		
							F26	No Health Consequences or Impact	A070101	Failure to Capture		
					CT3001	Implantable metal	E2403	No Clinical Signs, Symptoms or Conditions	A070101	Failure to Capture		
							F26	No Health Consequences or Impact	A070101	Failure to Capture		
					CT3002	Implantable synthetic polymer	E2403	No Clinical Signs, Symptoms or Conditions	A070101	Failure to Capture		
							F26	No Health Consequences or Impact	A070101	Failure to Capture		
					CT3006	Implantable inorganic material	E2403	No Clinical Signs, Symptoms or Conditions	A070101	Failure to Capture		
							F26	No Health Consequences or Impact	A070101	Failure to Capture		
					CT314	Electrically powered	E2403	No Clinical Signs, Symptoms or Conditions	A070101	Failure to Capture		
							F26	No Health Consequences or Impact	A070101	Failure to Capture		
					CT321	Long-term surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A070101	Failure to Capture		
							F26	No Health Consequences or Impact	A070101	Failure to Capture		
					CT449	Generators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070101	Failure to Capture		
							F26	No Health Consequences or Impact	A070101	Failure to Capture		
					CT981	Single use	E2403	No Clinical Signs, Symptoms or Conditions	A070101	Failure to Capture		
							F26	No Health Consequences or Impact	A070101	Failure to Capture		
					CT983	Surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A070101	Failure to Capture		
							F26	No Health Consequences or Impact	A070101	Failure to Capture		
6345	2022	All other reportable incidents	35223	Endocardial/interventricular septal pacing lead	CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070909	Oversensing		
									A072202	Low Impedance		
							F1908	Prolonged Surgery	A070909	Oversensing		
									A072202	Low Impedance		
6346	2022	All other reportable incidents	35852	Single-chamber implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	28/03/2022	28/03/2022
							F1901	Additional Surgery	A070504	Premature Discharge of Battery	28/03/2022	28/03/2022
6347	2022	All other reportable incidents	47265	Dual-chamber implantable pacemaker, rate-responsive	CT141	Electrodes/Leads and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070909	Oversensing		
							F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A070909	Oversensing		
					CT145	Cardiovascular devices	E2403	No Clinical Signs, Symptoms or Conditions	A070909	Oversensing		
							F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A070909	Oversensing		
					CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070909	Oversensing		
							F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A070909	Oversensing		
					CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A070909	Oversensing		
							F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A070909	Oversensing		
					CT233	Surgical	E2403	No Clinical Signs, Symptoms or Conditions	A070909	Oversensing		
							F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A070909	Oversensing		
					CT2406	Implantable	E2403	No Clinical Signs, Symptoms or Conditions	A070909	Oversensing		
							F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A070909	Oversensing		
					CT2407	Active implantable	E2403	No Clinical Signs, Symptoms or Conditions	A070909	Oversensing		
							F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A070909	Oversensing		

					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A070909	Oversensing		
							F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A070909	Oversensing		
					CT2727	Active implant devices/ancillaries	E2403	No Clinical Signs, Symptoms or Conditions	A070909	Oversensing		
							F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A070909	Oversensing		
					CT2750	Implantable cardiac electrical stimulators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070909	Oversensing		
							F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A070909	Oversensing		
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A070909	Oversensing		
							F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A070909	Oversensing		
					CT2985	Active	E2403	No Clinical Signs, Symptoms or Conditions	A070909	Oversensing		
							F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A070909	Oversensing		
					CT3001	Implantable metal	E2403	No Clinical Signs, Symptoms or Conditions	A070909	Oversensing		
							F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A070909	Oversensing		
					CT3002	Implantable synthetic polymer	E2403	No Clinical Signs, Symptoms or Conditions	A070909	Oversensing		
							F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A070909	Oversensing		
					CT3006	Implantable inorganic material	E2403	No Clinical Signs, Symptoms or Conditions	A070909	Oversensing		
							F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A070909	Oversensing		
					CT314	Electrically powered	E2403	No Clinical Signs, Symptoms or Conditions	A070909	Oversensing		
							F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A070909	Oversensing		
					CT321	Long-term surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A070909	Oversensing		
							F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A070909	Oversensing		
					CT449	Generators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070909	Oversensing		
							F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A070909	Oversensing		
					CT981	Single use	E2403	No Clinical Signs, Symptoms or Conditions	A070909	Oversensing		
							F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A070909	Oversensing		
					CT983	Surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A070909	Oversensing		
							F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A070909	Oversensing		
6348	2022	All other reportable incidents	17882	Physiologic-monitoring defibrillation system	CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A090811	Unable to Obtain Readings		
							F27	Problem Identified During Non-Clinical Procedure	A090811	Unable to Obtain Readings		
6349	2022	All other reportable incidents	47265	Dual-chamber implantable pacemaker, rate-responsive	CT141	Electrodes/Leads and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	21/03/2022	21/03/2022
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	21/03/2022	21/03/2022
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	21/03/2022	21/03/2022
					CT145	Cardiovascular devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	21/03/2022	21/03/2022
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	21/03/2022	21/03/2022
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	21/03/2022	21/03/2022
					CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	21/03/2022	21/03/2022
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	21/03/2022	21/03/2022
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	21/03/2022	21/03/2022
					CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	21/03/2022	21/03/2022
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	21/03/2022	21/03/2022
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	21/03/2022	21/03/2022
					CT233	Surgical	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	21/03/2022	21/03/2022
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	21/03/2022	21/03/2022
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	21/03/2022	21/03/2022
					CT2406	Implantable	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	21/03/2022	21/03/2022
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	21/03/2022	21/03/2022
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	21/03/2022	21/03/2022
					CT2407	Active implantable	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	21/03/2022	21/03/2022
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	21/03/2022	21/03/2022
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	21/03/2022	21/03/2022
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	21/03/2022	21/03/2022
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	21/03/2022	21/03/2022
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	21/03/2022	21/03/2022
					CT2727	Active implant devices/ancillaries	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	21/03/2022	21/03/2022
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	21/03/2022	21/03/2022
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	21/03/2022	21/03/2022
					CT2750	Implantable cardiac electrical stimulators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	21/03/2022	21/03/2022
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	21/03/2022	21/03/2022
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	21/03/2022	21/03/2022
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	21/03/2022	21/03/2022
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	21/03/2022	21/03/2022
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	21/03/2022	21/03/2022
					CT2985	Active	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	21/03/2022	21/03/2022
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	21/03/2022	21/03/2022
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	21/03/2022	21/03/2022
					CT3001	Implantable metal	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	21/03/2022	21/03/2022
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	21/03/2022	21/03/2022
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	21/03/2022	21/03/2022
					CT3002	Implantable synthetic polymer	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	21/03/2022	21/03/2022
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	21/03/2022	21/03/2022
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	21/03/2022	21/03/2022
					CT3006	Implantable inorganic material	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	21/03/2022	21/03/2022
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	21/03/2022	21/03/2022
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	21/03/2022	21/03/2022
					CT314	Electrically powered	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	21/03/2022	21/03/2022
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	21/03/2022	21/03/2022
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	21/03/2022	21/03/2022
					CT321	Long-term surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	21/03/2022	21/03/2022
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	21/03/2022	21/03/2022
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	21/03/2022	21/03/2022
					CT449	Generators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	21/03/2022	21/03/2022
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	21/03/2022	21/03/2022
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	21/03/2022	21/03/2022
					CT981	Single use	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	21/03/2022	21/03/2022
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	21/03/2022	21/03/2022
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	21/03/2022	21/03/2022
					CT983	Surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	21/03/2022	21/03/2022
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	21/03/2022	21/03/2022
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	21/03/2022	21/03/2022

6350	2022	All other reportable incidents	80789	Intracardiac pacemaker	CT130	Catheters and associated devices	E0605	Cardiac Tamponade	A24	No Apparent Device Problem		
							E0619	Pericardial Effusion	A24	No Apparent Device Problem		
							E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem		
							F08	Hospitalization or Prolonged Hospitalization	A24	No Apparent Device Problem		
							F12	Serious Injury/ Illness/ Impairment	A24	No Apparent Device Problem		
							F23	Unexpected Medical Intervention	A24	No Apparent Device Problem		
					CT145	Cardiovascular devices	E0605	Cardiac Tamponade	A24	No Apparent Device Problem		
							E0619	Pericardial Effusion	A24	No Apparent Device Problem		
							E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem		
							F08	Hospitalization or Prolonged Hospitalization	A24	No Apparent Device Problem		
							F12	Serious Injury/ Illness/ Impairment	A24	No Apparent Device Problem		
							F23	Unexpected Medical Intervention	A24	No Apparent Device Problem		
					CT1684	Cardiac pacemakers and associated devices	E0605	Cardiac Tamponade	A24	No Apparent Device Problem		
							E0619	Pericardial Effusion	A24	No Apparent Device Problem		
							E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem		
							F08	Hospitalization or Prolonged Hospitalization	A24	No Apparent Device Problem		
							F12	Serious Injury/ Illness/ Impairment	A24	No Apparent Device Problem		
							F23	Unexpected Medical Intervention	A24	No Apparent Device Problem		
					CT1685	Cardiac pacemakers	E0605	Cardiac Tamponade	A24	No Apparent Device Problem		
							E0619	Pericardial Effusion	A24	No Apparent Device Problem		
							E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem		
							F08	Hospitalization or Prolonged Hospitalization	A24	No Apparent Device Problem		
							F12	Serious Injury/ Illness/ Impairment	A24	No Apparent Device Problem		
							F23	Unexpected Medical Intervention	A24	No Apparent Device Problem		
					CT233	Surgical	E0605	Cardiac Tamponade	A24	No Apparent Device Problem		
							E0619	Pericardial Effusion	A24	No Apparent Device Problem		
							E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem		
							F08	Hospitalization or Prolonged Hospitalization	A24	No Apparent Device Problem		
							F12	Serious Injury/ Illness/ Impairment	A24	No Apparent Device Problem		
							F23	Unexpected Medical Intervention	A24	No Apparent Device Problem		
					CT2406	Implantable	E0605	Cardiac Tamponade	A24	No Apparent Device Problem		
							E0619	Pericardial Effusion	A24	No Apparent Device Problem		
							E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem		
							F08	Hospitalization or Prolonged Hospitalization	A24	No Apparent Device Problem		
							F12	Serious Injury/ Illness/ Impairment	A24	No Apparent Device Problem		
							F23	Unexpected Medical Intervention	A24	No Apparent Device Problem		
					CT2407	Active implantable	E0605	Cardiac Tamponade	A24	No Apparent Device Problem		
							E0619	Pericardial Effusion	A24	No Apparent Device Problem		
							E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem		
							F08	Hospitalization or Prolonged Hospitalization	A24	No Apparent Device Problem		
							F12	Serious Injury/ Illness/ Impairment	A24	No Apparent Device Problem		
							F23	Unexpected Medical Intervention	A24	No Apparent Device Problem		
					CT257	Implantable cardiac pacemakers	E0605	Cardiac Tamponade	A24	No Apparent Device Problem		
							E0619	Pericardial Effusion	A24	No Apparent Device Problem		
							E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem		
							F08	Hospitalization or Prolonged Hospitalization	A24	No Apparent Device Problem		
							F12	Serious Injury/ Illness/ Impairment	A24	No Apparent Device Problem		
							F23	Unexpected Medical Intervention	A24	No Apparent Device Problem		
					CT2727	Active implant devices/ancillaries	E0605	Cardiac Tamponade	A24	No Apparent Device Problem		
							E0619	Pericardial Effusion	A24	No Apparent Device Problem		
							E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem		
							F08	Hospitalization or Prolonged Hospitalization	A24	No Apparent Device Problem		
							F12	Serious Injury/ Illness/ Impairment	A24	No Apparent Device Problem		
							F23	Unexpected Medical Intervention	A24	No Apparent Device Problem		
					CT2750	Implantable cardiac electrical stimulators and associated devices	E0605	Cardiac Tamponade	A24	No Apparent Device Problem		
							E0619	Pericardial Effusion	A24	No Apparent Device Problem		
							E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem		
							F08	Hospitalization or Prolonged Hospitalization	A24	No Apparent Device Problem		
							F12	Serious Injury/ Illness/ Impairment	A24	No Apparent Device Problem		
							F23	Unexpected Medical Intervention	A24	No Apparent Device Problem		
					CT2818	Catheterization/Cannulation/Fluid path devices	E0605	Cardiac Tamponade	A24	No Apparent Device Problem		
							E0619	Pericardial Effusion	A24	No Apparent Device Problem		
							E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem		
							F08	Hospitalization or Prolonged Hospitalization	A24	No Apparent Device Problem		
							F12	Serious Injury/ Illness/ Impairment	A24	No Apparent Device Problem		
							F23	Unexpected Medical Intervention	A24	No Apparent Device Problem		
					CT2866	Implantable cardiac electrical stimulators	E0605	Cardiac Tamponade	A24	No Apparent Device Problem		
							E0619	Pericardial Effusion	A24	No Apparent Device Problem		
							E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem		
							F08	Hospitalization or Prolonged Hospitalization	A24	No Apparent Device Problem		
							F12	Serious Injury/ Illness/ Impairment	A24	No Apparent Device Problem		
							F23	Unexpected Medical Intervention	A24	No Apparent Device Problem		
					CT2985	Active	E0605	Cardiac Tamponade	A24	No Apparent Device Problem		
							E0619	Pericardial Effusion	A24	No Apparent Device Problem		
							E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem		
							F08	Hospitalization or Prolonged Hospitalization	A24	No Apparent Device Problem		
							F12	Serious Injury/ Illness/ Impairment	A24	No Apparent Device Problem		
							F23	Unexpected Medical Intervention	A24	No Apparent Device Problem		
					CT3001	Implantable metal	E0605	Cardiac Tamponade	A24	No Apparent Device Problem		
							E0619	Pericardial Effusion	A24	No Apparent Device Problem		
							E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem		
							F08	Hospitalization or Prolonged Hospitalization	A24	No Apparent Device Problem		
							F12	Serious Injury/ Illness/ Impairment	A24	No Apparent Device Problem		
							F23	Unexpected Medical Intervention	A24	No Apparent Device Problem		
					CT3002	Implantable synthetic polymer	E0605	Cardiac Tamponade	A24	No Apparent Device Problem		
							E0619	Pericardial Effusion	A24	No Apparent Device Problem		
							E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem		
							F08	Hospitalization or Prolonged Hospitalization	A24	No Apparent Device Problem		

							F12	Serious Injury/ Illness/ Impairment	A24	No Apparent Device Problem		
							F23	Unexpected Medical Intervention	A24	No Apparent Device Problem		
					CT3006	Implantable inorganic material	E0605	Cardiac Tamponade	A24	No Apparent Device Problem		
							E0619	Pericardial Effusion	A24	No Apparent Device Problem		
							E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem		
							F08	Hospitalization or Prolonged Hospitalization	A24	No Apparent Device Problem		
							F12	Serious Injury/ Illness/ Impairment	A24	No Apparent Device Problem		
							F23	Unexpected Medical Intervention	A24	No Apparent Device Problem		
					CT314	Electrically powered	E0605	Cardiac Tamponade	A24	No Apparent Device Problem		
							E0619	Pericardial Effusion	A24	No Apparent Device Problem		
							E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem		
							F08	Hospitalization or Prolonged Hospitalization	A24	No Apparent Device Problem		
							F12	Serious Injury/ Illness/ Impairment	A24	No Apparent Device Problem		
							F23	Unexpected Medical Intervention	A24	No Apparent Device Problem		
					CT321	Long-term surgical invasive	E0605	Cardiac Tamponade	A24	No Apparent Device Problem		
							E0619	Pericardial Effusion	A24	No Apparent Device Problem		
							E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem		
							F08	Hospitalization or Prolonged Hospitalization	A24	No Apparent Device Problem		
							F12	Serious Injury/ Illness/ Impairment	A24	No Apparent Device Problem		
							F23	Unexpected Medical Intervention	A24	No Apparent Device Problem		
					CT981	Single use	E0605	Cardiac Tamponade	A24	No Apparent Device Problem		
							E0619	Pericardial Effusion	A24	No Apparent Device Problem		
							E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem		
							F08	Hospitalization or Prolonged Hospitalization	A24	No Apparent Device Problem		
							F12	Serious Injury/ Illness/ Impairment	A24	No Apparent Device Problem		
							F23	Unexpected Medical Intervention	A24	No Apparent Device Problem		
					CT983	Surgical invasive	E0605	Cardiac Tamponade	A24	No Apparent Device Problem		
							E0619	Pericardial Effusion	A24	No Apparent Device Problem		
							E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem		
							F08	Hospitalization or Prolonged Hospitalization	A24	No Apparent Device Problem		
							F12	Serious Injury/ Illness/ Impairment	A24	No Apparent Device Problem		
							F23	Unexpected Medical Intervention	A24	No Apparent Device Problem		
6351	2021	All other reportable incidents	47265	Dual-chamber implantable pacemaker, rate-responsive	CT141	Electrodes/Leads and associated devices	E2007	Fall	A0401	Break	11/08/2021	11/08/2021
									A05	Mechanical Problem	11/08/2021	11/08/2021
									A070909	Oversensing	11/08/2021	11/08/2021
									A090801	Signal Artifact/Noise	11/08/2021	11/08/2021
							E2403	No Clinical Signs, Symptoms or Conditions	A0401	Break	11/08/2021	11/08/2021
									A05	Mechanical Problem	11/08/2021	11/08/2021
									A070909	Oversensing	11/08/2021	11/08/2021
									A090801	Signal Artifact/Noise	11/08/2021	11/08/2021
							F11	Non-Serious Injury/ Illness/ Impairment	A0401	Break	11/08/2021	11/08/2021
									A05	Mechanical Problem	11/08/2021	11/08/2021
									A070909	Oversensing	11/08/2021	11/08/2021
									A090801	Signal Artifact/Noise	11/08/2021	11/08/2021
							F1903	Device Explantation	A0401	Break	11/08/2021	11/08/2021
									A05	Mechanical Problem	11/08/2021	11/08/2021
									A070909	Oversensing	11/08/2021	11/08/2021
									A090801	Signal Artifact/Noise	11/08/2021	11/08/2021
					CT145	Cardiovascular devices	E2007	Fall	A0401	Break	11/08/2021	11/08/2021
									A05	Mechanical Problem	11/08/2021	11/08/2021
									A070909	Oversensing	11/08/2021	11/08/2021
									A090801	Signal Artifact/Noise	11/08/2021	11/08/2021
							E2403	No Clinical Signs, Symptoms or Conditions	A0401	Break	11/08/2021	11/08/2021
									A05	Mechanical Problem	11/08/2021	11/08/2021
									A070909	Oversensing	11/08/2021	11/08/2021
									A090801	Signal Artifact/Noise	11/08/2021	11/08/2021
							F11	Non-Serious Injury/ Illness/ Impairment	A0401	Break	11/08/2021	11/08/2021
									A05	Mechanical Problem	11/08/2021	11/08/2021
									A070909	Oversensing	11/08/2021	11/08/2021
									A090801	Signal Artifact/Noise	11/08/2021	11/08/2021
							F1903	Device Explantation	A0401	Break	11/08/2021	11/08/2021
									A05	Mechanical Problem	11/08/2021	11/08/2021
									A070909	Oversensing	11/08/2021	11/08/2021
									A090801	Signal Artifact/Noise	11/08/2021	11/08/2021
					CT1684	Cardiac pacemakers and associated devices	E2007	Fall	A0401	Break	11/08/2021	11/08/2021
									A05	Mechanical Problem	11/08/2021	11/08/2021
									A070909	Oversensing	11/08/2021	11/08/2021
									A090801	Signal Artifact/Noise	11/08/2021	11/08/2021
							E2403	No Clinical Signs, Symptoms or Conditions	A0401	Break	11/08/2021	11/08/2021
									A05	Mechanical Problem	11/08/2021	11/08/2021
									A070909	Oversensing	11/08/2021	11/08/2021
									A090801	Signal Artifact/Noise	11/08/2021	11/08/2021
							F11	Non-Serious Injury/ Illness/ Impairment	A0401	Break	11/08/2021	11/08/2021
									A05	Mechanical Problem	11/08/2021	11/08/2021
									A070909	Oversensing	11/08/2021	11/08/2021
									A090801	Signal Artifact/Noise	11/08/2021	11/08/2021
							F1903	Device Explantation	A0401	Break	11/08/2021	11/08/2021
									A05	Mechanical Problem	11/08/2021	11/08/2021
									A070909	Oversensing	11/08/2021	11/08/2021
									A090801	Signal Artifact/Noise	11/08/2021	11/08/2021
					CT1685	Cardiac pacemakers	E2007	Fall	A0401	Break	11/08/2021	11/08/2021
									A05	Mechanical Problem	11/08/2021	11/08/2021
									A070909	Oversensing	11/08/2021	11/08/2021
									A090801	Signal Artifact/Noise	11/08/2021	11/08/2021
							E2403	No Clinical Signs, Symptoms or Conditions	A0401	Break	11/08/2021	11/08/2021
									A05	Mechanical Problem	11/08/2021	11/08/2021
									A070909	Oversensing	11/08/2021	11/08/2021
									A090801	Signal Artifact/Noise	11/08/2021	11/08/2021

						F11	Non-Serious Injury/ Illness/ Impairment	A0401	Break	11/08/2021	11/08/2021
								A05	Mechanical Problem	11/08/2021	11/08/2021
								A070909	Oversensing	11/08/2021	11/08/2021
								A090801	Signal Artifact/Noise	11/08/2021	11/08/2021
						F1903	Device Explantation	A0401	Break	11/08/2021	11/08/2021
								A05	Mechanical Problem	11/08/2021	11/08/2021
								A070909	Oversensing	11/08/2021	11/08/2021
								A090801	Signal Artifact/Noise	11/08/2021	11/08/2021
					CT233	Surgical	E2007	Fall	A0401	Break	11/08/2021	11/08/2021
									A05	Mechanical Problem	11/08/2021	11/08/2021
									A070909	Oversensing	11/08/2021	11/08/2021
									A090801	Signal Artifact/Noise	11/08/2021	11/08/2021
							E2403	No Clinical Signs, Symptoms or Conditions	A0401	Break	11/08/2021	11/08/2021
									A05	Mechanical Problem	11/08/2021	11/08/2021
									A070909	Oversensing	11/08/2021	11/08/2021
									A090801	Signal Artifact/Noise	11/08/2021	11/08/2021
							F11	Non-Serious Injury/ Illness/ Impairment	A0401	Break	11/08/2021	11/08/2021
									A05	Mechanical Problem	11/08/2021	11/08/2021
									A070909	Oversensing	11/08/2021	11/08/2021
									A090801	Signal Artifact/Noise	11/08/2021	11/08/2021
							F1903	Device Explantation	A0401	Break	11/08/2021	11/08/2021
									A05	Mechanical Problem	11/08/2021	11/08/2021
									A070909	Oversensing	11/08/2021	11/08/2021
									A090801	Signal Artifact/Noise	11/08/2021	11/08/2021
					CT2406	Implantable	E2007	Fall	A0401	Break	11/08/2021	11/08/2021
									A05	Mechanical Problem	11/08/2021	11/08/2021
									A070909	Oversensing	11/08/2021	11/08/2021
									A090801	Signal Artifact/Noise	11/08/2021	11/08/2021
							E2403	No Clinical Signs, Symptoms or Conditions	A0401	Break	11/08/2021	11/08/2021
									A05	Mechanical Problem	11/08/2021	11/08/2021
									A070909	Oversensing	11/08/2021	11/08/2021
									A090801	Signal Artifact/Noise	11/08/2021	11/08/2021
							F11	Non-Serious Injury/ Illness/ Impairment	A0401	Break	11/08/2021	11/08/2021
									A05	Mechanical Problem	11/08/2021	11/08/2021
									A070909	Oversensing	11/08/2021	11/08/2021
									A090801	Signal Artifact/Noise	11/08/2021	11/08/2021
							F1903	Device Explantation	A0401	Break	11/08/2021	11/08/2021
									A05	Mechanical Problem	11/08/2021	11/08/2021
									A070909	Oversensing	11/08/2021	11/08/2021
									A090801	Signal Artifact/Noise	11/08/2021	11/08/2021
					CT2407	Active implantable	E2007	Fall	A0401	Break	11/08/2021	11/08/2021
									A05	Mechanical Problem	11/08/2021	11/08/2021
									A070909	Oversensing	11/08/2021	11/08/2021
									A090801	Signal Artifact/Noise	11/08/2021	11/08/2021
							E2403	No Clinical Signs, Symptoms or Conditions	A0401	Break	11/08/2021	11/08/2021
									A05	Mechanical Problem	11/08/2021	11/08/2021
									A070909	Oversensing	11/08/2021	11/08/2021
									A090801	Signal Artifact/Noise	11/08/2021	11/08/2021
							F11	Non-Serious Injury/ Illness/ Impairment	A0401	Break	11/08/2021	11/08/2021
									A05	Mechanical Problem	11/08/2021	11/08/2021
									A070909	Oversensing	11/08/2021	11/08/2021
									A090801	Signal Artifact/Noise	11/08/2021	11/08/2021
							F1903	Device Explantation	A0401	Break	11/08/2021	11/08/2021
									A05	Mechanical Problem	11/08/2021	11/08/2021
									A070909	Oversensing	11/08/2021	11/08/2021
									A090801	Signal Artifact/Noise	11/08/2021	11/08/2021
					CT257	Implantable cardiac pacemakers	E2007	Fall	A0401	Break	11/08/2021	11/08/2021
									A05	Mechanical Problem	11/08/2021	11/08/2021
									A070909	Oversensing	11/08/2021	11/08/2021
									A090801	Signal Artifact/Noise	11/08/2021	11/08/2021
							E2403	No Clinical Signs, Symptoms or Conditions	A0401	Break	11/08/2021	11/08/2021
									A05	Mechanical Problem	11/08/2021	11/08/2021
									A070909	Oversensing	11/08/2021	11/08/2021
									A090801	Signal Artifact/Noise	11/08/2021	11/08/2021
							F11	Non-Serious Injury/ Illness/ Impairment	A0401	Break	11/08/2021	11/08/2021
									A05	Mechanical Problem	11/08/2021	11/08/2021
									A070909	Oversensing	11/08/2021	11/08/2021
									A090801	Signal Artifact/Noise	11/08/2021	11/08/2021
							F1903	Device Explantation	A0401	Break	11/08/2021	11/08/2021
									A05	Mechanical Problem	11/08/2021	11/08/2021
									A070909	Oversensing	11/08/2021	11/08/2021
									A090801	Signal Artifact/Noise	11/08/2021	11/08/2021
					CT2727	Active implant devices/ancillaries	E2007	Fall	A0401	Break	11/08/2021	11/08/2021
									A05	Mechanical Problem	11/08/2021	11/08/2021
									A070909	Oversensing	11/08/2021	11/08/2021
									A090801	Signal Artifact/Noise	11/08/2021	11/08/2021
							E2403	No Clinical Signs, Symptoms or Conditions	A0401	Break	11/08/2021	11/08/2021
									A05	Mechanical Problem	11/08/2021	11/08/2021
									A070909	Oversensing	11/08/2021	11/08/2021
									A090801	Signal Artifact/Noise	11/08/2021	11/08/2021
							F11	Non-Serious Injury/ Illness/ Impairment	A0401	Break	11/08/2021	11/08/2021
									A05	Mechanical Problem	11/08/2021	11/08/2021
									A070909	Oversensing	11/08/2021	11/08/2021
									A090801	Signal Artifact/Noise	11/08/2021	11/08/2021
							F1903	Device Explantation	A0401	Break	11/08/2021	11/08/2021
									A05	Mechanical Problem	11/08/2021	11/08/2021
									A070909	Oversensing	11/08/2021	11/08/2021
									A090801	Signal Artifact/Noise	11/08/2021	11/08/2021

					CT2750	Implantable cardiac electrical stimulators and associated devices	E2007	Fall	A0401	Break	11/08/2021	11/08/2021
									A05	Mechanical Problem	11/08/2021	11/08/2021
									A070909	Oversensing	11/08/2021	11/08/2021
									A090801	Signal Artifact/Noise	11/08/2021	11/08/2021
							E2403	No Clinical Signs, Symptoms or Conditions	A0401	Break	11/08/2021	11/08/2021
									A05	Mechanical Problem	11/08/2021	11/08/2021
									A070909	Oversensing	11/08/2021	11/08/2021
									A090801	Signal Artifact/Noise	11/08/2021	11/08/2021
							F11	Non-Serious Injury/ Illness/ Impairment	A0401	Break	11/08/2021	11/08/2021
									A05	Mechanical Problem	11/08/2021	11/08/2021
									A070909	Oversensing	11/08/2021	11/08/2021
									A090801	Signal Artifact/Noise	11/08/2021	11/08/2021
							F1903	Device Explantation	A0401	Break	11/08/2021	11/08/2021
									A05	Mechanical Problem	11/08/2021	11/08/2021
									A070909	Oversensing	11/08/2021	11/08/2021
									A090801	Signal Artifact/Noise	11/08/2021	11/08/2021
					CT2866	Implantable cardiac electrical stimulators	E2007	Fall	A0401	Break	11/08/2021	11/08/2021
									A05	Mechanical Problem	11/08/2021	11/08/2021
									A070909	Oversensing	11/08/2021	11/08/2021
									A090801	Signal Artifact/Noise	11/08/2021	11/08/2021
							E2403	No Clinical Signs, Symptoms or Conditions	A0401	Break	11/08/2021	11/08/2021
									A05	Mechanical Problem	11/08/2021	11/08/2021
									A070909	Oversensing	11/08/2021	11/08/2021
									A090801	Signal Artifact/Noise	11/08/2021	11/08/2021
							F11	Non-Serious Injury/ Illness/ Impairment	A0401	Break	11/08/2021	11/08/2021
									A05	Mechanical Problem	11/08/2021	11/08/2021
									A070909	Oversensing	11/08/2021	11/08/2021
									A090801	Signal Artifact/Noise	11/08/2021	11/08/2021
							F1903	Device Explantation	A0401	Break	11/08/2021	11/08/2021
									A05	Mechanical Problem	11/08/2021	11/08/2021
									A070909	Oversensing	11/08/2021	11/08/2021
									A090801	Signal Artifact/Noise	11/08/2021	11/08/2021
					CT2985	Active	E2007	Fall	A0401	Break	11/08/2021	11/08/2021
									A05	Mechanical Problem	11/08/2021	11/08/2021
									A070909	Oversensing	11/08/2021	11/08/2021
									A090801	Signal Artifact/Noise	11/08/2021	11/08/2021
							E2403	No Clinical Signs, Symptoms or Conditions	A0401	Break	11/08/2021	11/08/2021
									A05	Mechanical Problem	11/08/2021	11/08/2021
									A070909	Oversensing	11/08/2021	11/08/2021
									A090801	Signal Artifact/Noise	11/08/2021	11/08/2021
							F11	Non-Serious Injury/ Illness/ Impairment	A0401	Break	11/08/2021	11/08/2021
									A05	Mechanical Problem	11/08/2021	11/08/2021
									A070909	Oversensing	11/08/2021	11/08/2021
									A090801	Signal Artifact/Noise	11/08/2021	11/08/2021
							F1903	Device Explantation	A0401	Break	11/08/2021	11/08/2021
									A05	Mechanical Problem	11/08/2021	11/08/2021
									A070909	Oversensing	11/08/2021	11/08/2021
									A090801	Signal Artifact/Noise	11/08/2021	11/08/2021
					CT3001	Implantable metal	E2007	Fall	A0401	Break	11/08/2021	11/08/2021
									A05	Mechanical Problem	11/08/2021	11/08/2021
									A070909	Oversensing	11/08/2021	11/08/2021
									A090801	Signal Artifact/Noise	11/08/2021	11/08/2021
							E2403	No Clinical Signs, Symptoms or Conditions	A0401	Break	11/08/2021	11/08/2021
									A05	Mechanical Problem	11/08/2021	11/08/2021
									A070909	Oversensing	11/08/2021	11/08/2021
									A090801	Signal Artifact/Noise	11/08/2021	11/08/2021
							F11	Non-Serious Injury/ Illness/ Impairment	A0401	Break	11/08/2021	11/08/2021
									A05	Mechanical Problem	11/08/2021	11/08/2021
									A070909	Oversensing	11/08/2021	11/08/2021
									A090801	Signal Artifact/Noise	11/08/2021	11/08/2021
							F1903	Device Explantation	A0401	Break	11/08/2021	11/08/2021
									A05	Mechanical Problem	11/08/2021	11/08/2021
									A070909	Oversensing	11/08/2021	11/08/2021
									A090801	Signal Artifact/Noise	11/08/2021	11/08/2021
					CT3002	Implantable synthetic polymer	E2007	Fall	A0401	Break	11/08/2021	11/08/2021
									A05	Mechanical Problem	11/08/2021	11/08/2021
									A070909	Oversensing	11/08/2021	11/08/2021
									A090801	Signal Artifact/Noise	11/08/2021	11/08/2021
							E2403	No Clinical Signs, Symptoms or Conditions	A0401	Break	11/08/2021	11/08/2021
									A05	Mechanical Problem	11/08/2021	11/08/2021
									A070909	Oversensing	11/08/2021	11/08/2021
									A090801	Signal Artifact/Noise	11/08/2021	11/08/2021
							F11	Non-Serious Injury/ Illness/ Impairment	A0401	Break	11/08/2021	11/08/2021
									A05	Mechanical Problem	11/08/2021	11/08/2021
									A070909	Oversensing	11/08/2021	11/08/2021
									A090801	Signal Artifact/Noise	11/08/2021	11/08/2021
							F1903	Device Explantation	A0401	Break	11/08/2021	11/08/2021
									A05	Mechanical Problem	11/08/2021	11/08/2021
									A070909	Oversensing	11/08/2021	11/08/2021
									A090801	Signal Artifact/Noise	11/08/2021	11/08/2021
					CT3006	Implantable inorganic material	E2007	Fall	A0401	Break	11/08/2021	11/08/2021
									A05	Mechanical Problem	11/08/2021	11/08/2021
									A070909	Oversensing	11/08/2021	11/08/2021
									A090801	Signal Artifact/Noise	11/08/2021	11/08/2021
							E2403	No Clinical Signs, Symptoms or Conditions	A0401	Break	11/08/2021	11/08/2021
									A05	Mechanical Problem	11/08/2021	11/08/2021
									A070909	Oversensing	11/08/2021	11/08/2021
									A090801	Signal Artifact/Noise	11/08/2021	11/08/2021

							F11	Non-Serious Injury/ Illness/ Impairment	A0401	Break	11/08/2021	11/08/2021
									A05	Mechanical Problem	11/08/2021	11/08/2021
									A070909	Oversensing	11/08/2021	11/08/2021
									A090801	Signal Artifact/Noise	11/08/2021	11/08/2021
							F1903	Device Explantation	A0401	Break	11/08/2021	11/08/2021
									A05	Mechanical Problem	11/08/2021	11/08/2021
									A070909	Oversensing	11/08/2021	11/08/2021
									A090801	Signal Artifact/Noise	11/08/2021	11/08/2021
					CT314	Electrically powered	E2007	Fall	A0401	Break	11/08/2021	11/08/2021
									A05	Mechanical Problem	11/08/2021	11/08/2021
									A070909	Oversensing	11/08/2021	11/08/2021
									A090801	Signal Artifact/Noise	11/08/2021	11/08/2021
							E2403	No Clinical Signs, Symptoms or Conditions	A0401	Break	11/08/2021	11/08/2021
									A05	Mechanical Problem	11/08/2021	11/08/2021
									A070909	Oversensing	11/08/2021	11/08/2021
									A090801	Signal Artifact/Noise	11/08/2021	11/08/2021
							F11	Non-Serious Injury/ Illness/ Impairment	A0401	Break	11/08/2021	11/08/2021
									A05	Mechanical Problem	11/08/2021	11/08/2021
									A070909	Oversensing	11/08/2021	11/08/2021
									A090801	Signal Artifact/Noise	11/08/2021	11/08/2021
							F1903	Device Explantation	A0401	Break	11/08/2021	11/08/2021
									A05	Mechanical Problem	11/08/2021	11/08/2021
									A070909	Oversensing	11/08/2021	11/08/2021
									A090801	Signal Artifact/Noise	11/08/2021	11/08/2021
					CT321	Long-term surgical invasive	E2007	Fall	A0401	Break	11/08/2021	11/08/2021
									A05	Mechanical Problem	11/08/2021	11/08/2021
									A070909	Oversensing	11/08/2021	11/08/2021
									A090801	Signal Artifact/Noise	11/08/2021	11/08/2021
							E2403	No Clinical Signs, Symptoms or Conditions	A0401	Break	11/08/2021	11/08/2021
									A05	Mechanical Problem	11/08/2021	11/08/2021
									A070909	Oversensing	11/08/2021	11/08/2021
									A090801	Signal Artifact/Noise	11/08/2021	11/08/2021
							F11	Non-Serious Injury/ Illness/ Impairment	A0401	Break	11/08/2021	11/08/2021
									A05	Mechanical Problem	11/08/2021	11/08/2021
									A070909	Oversensing	11/08/2021	11/08/2021
									A090801	Signal Artifact/Noise	11/08/2021	11/08/2021
							F1903	Device Explantation	A0401	Break	11/08/2021	11/08/2021
									A05	Mechanical Problem	11/08/2021	11/08/2021
									A070909	Oversensing	11/08/2021	11/08/2021
									A090801	Signal Artifact/Noise	11/08/2021	11/08/2021
					CT449	Generators and associated devices	E2007	Fall	A0401	Break	11/08/2021	11/08/2021
									A05	Mechanical Problem	11/08/2021	11/08/2021
									A070909	Oversensing	11/08/2021	11/08/2021
									A090801	Signal Artifact/Noise	11/08/2021	11/08/2021
							E2403	No Clinical Signs, Symptoms or Conditions	A0401	Break	11/08/2021	11/08/2021
									A05	Mechanical Problem	11/08/2021	11/08/2021
									A070909	Oversensing	11/08/2021	11/08/2021
									A090801	Signal Artifact/Noise	11/08/2021	11/08/2021
							F11	Non-Serious Injury/ Illness/ Impairment	A0401	Break	11/08/2021	11/08/2021
									A05	Mechanical Problem	11/08/2021	11/08/2021
									A070909	Oversensing	11/08/2021	11/08/2021
									A090801	Signal Artifact/Noise	11/08/2021	11/08/2021
							F1903	Device Explantation	A0401	Break	11/08/2021	11/08/2021
									A05	Mechanical Problem	11/08/2021	11/08/2021
									A070909	Oversensing	11/08/2021	11/08/2021
									A090801	Signal Artifact/Noise	11/08/2021	11/08/2021
					CT981	Single use	E2007	Fall	A0401	Break	11/08/2021	11/08/2021
									A05	Mechanical Problem	11/08/2021	11/08/2021
									A070909	Oversensing	11/08/2021	11/08/2021
									A090801	Signal Artifact/Noise	11/08/2021	11/08/2021
							E2403	No Clinical Signs, Symptoms or Conditions	A0401	Break	11/08/2021	11/08/2021
									A05	Mechanical Problem	11/08/2021	11/08/2021
									A070909	Oversensing	11/08/2021	11/08/2021
									A090801	Signal Artifact/Noise	11/08/2021	11/08/2021
							F11	Non-Serious Injury/ Illness/ Impairment	A0401	Break	11/08/2021	11/08/2021
									A05	Mechanical Problem	11/08/2021	11/08/2021
									A070909	Oversensing	11/08/2021	11/08/2021
									A090801	Signal Artifact/Noise	11/08/2021	11/08/2021
							F1903	Device Explantation	A0401	Break	11/08/2021	11/08/2021
									A05	Mechanical Problem	11/08/2021	11/08/2021
									A070909	Oversensing	11/08/2021	11/08/2021
									A090801	Signal Artifact/Noise	11/08/2021	11/08/2021
					CT983	Surgical invasive	E2007	Fall	A0401	Break	11/08/2021	11/08/2021
									A05	Mechanical Problem	11/08/2021	11/08/2021
									A070909	Oversensing	11/08/2021	11/08/2021
									A090801	Signal Artifact/Noise	11/08/2021	11/08/2021
							E2403	No Clinical Signs, Symptoms or Conditions	A0401	Break	11/08/2021	11/08/2021
									A05	Mechanical Problem	11/08/2021	11/08/2021
									A070909	Oversensing	11/08/2021	11/08/2021
									A090801	Signal Artifact/Noise	11/08/2021	11/08/2021
							F11	Non-Serious Injury/ Illness/ Impairment	A0401	Break	11/08/2021	11/08/2021
									A05	Mechanical Problem	11/08/2021	11/08/2021
									A070909	Oversensing	11/08/2021	11/08/2021
									A090801	Signal Artifact/Noise	11/08/2021	11/08/2021
							F1903	Device Explantation	A0401	Break	11/08/2021	11/08/2021
									A05	Mechanical Problem	11/08/2021	11/08/2021
									A070909	Oversensing	11/08/2021	11/08/2021
									A090801	Signal Artifact/Noise	11/08/2021	11/08/2021

6352	2021	Death	35223	Endocardial/interventricular septal pacing lead	CT1684	Cardiac pacemakers and associated devices	E0602	Cardiac Arrest	A24	No Apparent Device Problem		
							F02	Death	A24	No Apparent Device Problem		
6353	2021	All other reportable incidents	47263	Cardiac resynchronization therapy implantable pacemaker	CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A0908	Incorrect, Inadequate or Imprecise Result or Readings	04/08/2021	04/08/2021
							F08	Hospitalization or Prolonged Hospitalization	A0908	Incorrect, Inadequate or Imprecise Result or Readings	04/08/2021	04/08/2021
							F1905	Device Revision or Replacement	A0908	Incorrect, Inadequate or Imprecise Result or Readings	04/08/2021	04/08/2021
					CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A0908	Incorrect, Inadequate or Imprecise Result or Readings	04/08/2021	04/08/2021
							F08	Hospitalization or Prolonged Hospitalization	A0908	Incorrect, Inadequate or Imprecise Result or Readings	04/08/2021	04/08/2021
							F1905	Device Revision or Replacement	A0908	Incorrect, Inadequate or Imprecise Result or Readings	04/08/2021	04/08/2021
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A0908	Incorrect, Inadequate or Imprecise Result or Readings	04/08/2021	04/08/2021
							F08	Hospitalization or Prolonged Hospitalization	A0908	Incorrect, Inadequate or Imprecise Result or Readings	04/08/2021	04/08/2021
							F1905	Device Revision or Replacement	A0908	Incorrect, Inadequate or Imprecise Result or Readings	04/08/2021	04/08/2021
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A0908	Incorrect, Inadequate or Imprecise Result or Readings	04/08/2021	04/08/2021
							F08	Hospitalization or Prolonged Hospitalization	A0908	Incorrect, Inadequate or Imprecise Result or Readings	04/08/2021	04/08/2021
							F1905	Device Revision or Replacement	A0908	Incorrect, Inadequate or Imprecise Result or Readings	04/08/2021	04/08/2021
6354	2021	All other reportable incidents	17882	Physiologic-monitoring defibrillation system	CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A0710	Failure to Conduct		
							F27	Problem Identified During Non-Clinical Procedure	A0710	Failure to Conduct		
6355	2022	All other reportable incidents	35223	Endocardial/interventricular septal pacing lead	CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070909	Oversensing	30/03/2022	30/03/2022
							F1905	Device Revision or Replacement	A070909	Oversensing	30/03/2022	30/03/2022
6356	2021	All other reportable incidents	47265	Dual-chamber implantable pacemaker, rate-responsive	CT141	Electrodes/Leads and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT145	Cardiovascular devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT233	Surgical	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT2406	Implantable	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT2407	Active implantable	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT2727	Active implant devices/ancillaries	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT2750	Implantable cardiac electrical stimulators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT2985	Active	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT3001	Implantable metal	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT3002	Implantable synthetic polymer	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT3006	Implantable inorganic material	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT314	Electrically powered	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT321	Long-term surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT449	Generators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT981	Single use	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT983	Surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
6357	2022	All other reportable incidents	17882	Physiologic-monitoring defibrillation system	CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070803	Failure to Power Up		
							F26	No Health Consequences or Impact	A070803	Failure to Power Up		
6358	2022	All other reportable incidents	47265	Dual-chamber implantable pacemaker, rate-responsive	CT141	Electrodes/Leads and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT145	Cardiovascular devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT233	Surgical	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT2406	Implantable	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT2407	Active implantable	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT2727	Active implant devices/ancillaries	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT2750	Implantable cardiac electrical stimulators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		

							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT2985	Active	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT3001	Implantable metal	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT3002	Implantable synthetic polymer	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT3006	Implantable inorganic material	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT314	Electrically powered	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT321	Long-term surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT449	Generators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT981	Single use	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT983	Surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
6359	2022	All other reportable incidents	35223	Endocardial/interventricular septal pacing lead	CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A0722	Impedance Problem		
							F26	No Health Consequences or Impact	A0722	Impedance Problem		
6360	2022	All other reportable incidents	35223	Endocardial/interventricular septal pacing lead	CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070101	Failure to Capture	29/07/2021	29/07/2021
							F08	Hospitalization or Prolonged Hospitalization	A070101	Failure to Capture	29/07/2021	29/07/2021
							F1905	Device Revision or Replacement	A070101	Failure to Capture	29/07/2021	29/07/2021
6361	2021	All other reportable incidents	47270	Cardiac resynchronization therapy implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E0601	Arrhythmia	A070910	Undersensing		
									A0908	Incorrect, Inadequate or Imprecise Result or Readings		
							E2403	No Clinical Signs, Symptoms or Conditions	A070910	Undersensing		
									A0908	Incorrect, Inadequate or Imprecise Result or Readings		
							F02	Death	A070910	Undersensing		
									A0908	Incorrect, Inadequate or Imprecise Result or Readings		
6362	2021	All other reportable incidents	17882	Physiologic-monitoring defibrillation system	CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A0708	Power Problem		
							F27	Problem Identified During Non-Clinical Procedure	A0708	Power Problem		
6363	2021	All other reportable incidents	47265	Dual-chamber implantable pacemaker, rate-responsive	CT141	Electrodes/Leads and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT145	Cardiovascular devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT233	Surgical	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT2406	Implantable	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT2407	Active implantable	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT2727	Active implant devices/ancillaries	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT2750	Implantable cardiac electrical stimulators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT2985	Active	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT3001	Implantable metal	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT3002	Implantable synthetic polymer	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT3006	Implantable inorganic material	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT314	Electrically powered	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT321	Long-term surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT449	Generators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT981	Single use	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT983	Surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
6364	2022	All other reportable incidents	47267	Single-chamber implantable pacemaker, rate-responsive	CT141	Electrodes/Leads and associated devices	E0112	Dizziness	A09	Output Problem		
							E2403	No Clinical Signs, Symptoms or Conditions	A09	Output Problem		
							F08	Hospitalization or Prolonged Hospitalization	A09	Output Problem		
							F11	Non-Serious Injury/ Illness/ Impairment	A09	Output Problem		
							F1905	Device Revision or Replacement	A09	Output Problem		
					CT145	Cardiovascular devices	E0112	Dizziness	A09	Output Problem		
							E2403	No Clinical Signs, Symptoms or Conditions	A09	Output Problem		
							F08	Hospitalization or Prolonged Hospitalization	A09	Output Problem		
							F11	Non-Serious Injury/ Illness/ Impairment	A09	Output Problem		
							F1905	Device Revision or Replacement	A09	Output Problem		
					CT1684	Cardiac pacemakers and associated devices	E0112	Dizziness	A09	Output Problem		

							E2403	No Clinical Signs, Symptoms or Conditions	A09	Output Problem		
							F08	Hospitalization or Prolonged Hospitalization	A09	Output Problem		
							F11	Non-Serious Injury/ Illness/ Impairment	A09	Output Problem		
							F1905	Device Revision or Replacement	A09	Output Problem		
					CT1685	Cardiac pacemakers	E0112	Dizziness	A09	Output Problem		
							E2403	No Clinical Signs, Symptoms or Conditions	A09	Output Problem		
							F08	Hospitalization or Prolonged Hospitalization	A09	Output Problem		
							F11	Non-Serious Injury/ Illness/ Impairment	A09	Output Problem		
							F1905	Device Revision or Replacement	A09	Output Problem		
					CT233	Surgical	E0112	Dizziness	A09	Output Problem		
							E2403	No Clinical Signs, Symptoms or Conditions	A09	Output Problem		
							F08	Hospitalization or Prolonged Hospitalization	A09	Output Problem		
							F11	Non-Serious Injury/ Illness/ Impairment	A09	Output Problem		
							F1905	Device Revision or Replacement	A09	Output Problem		
					CT2406	Implantable	E0112	Dizziness	A09	Output Problem		
							E2403	No Clinical Signs, Symptoms or Conditions	A09	Output Problem		
							F08	Hospitalization or Prolonged Hospitalization	A09	Output Problem		
							F11	Non-Serious Injury/ Illness/ Impairment	A09	Output Problem		
							F1905	Device Revision or Replacement	A09	Output Problem		
					CT2407	Active implantable	E0112	Dizziness	A09	Output Problem		
							E2403	No Clinical Signs, Symptoms or Conditions	A09	Output Problem		
							F08	Hospitalization or Prolonged Hospitalization	A09	Output Problem		
							F11	Non-Serious Injury/ Illness/ Impairment	A09	Output Problem		
							F1905	Device Revision or Replacement	A09	Output Problem		
					CT257	Implantable cardiac pacemakers	E0112	Dizziness	A09	Output Problem		
							E2403	No Clinical Signs, Symptoms or Conditions	A09	Output Problem		
							F08	Hospitalization or Prolonged Hospitalization	A09	Output Problem		
							F11	Non-Serious Injury/ Illness/ Impairment	A09	Output Problem		
							F1905	Device Revision or Replacement	A09	Output Problem		
					CT2727	Active implant devices/ancillaries	E0112	Dizziness	A09	Output Problem		
							E2403	No Clinical Signs, Symptoms or Conditions	A09	Output Problem		
							F08	Hospitalization or Prolonged Hospitalization	A09	Output Problem		
							F11	Non-Serious Injury/ Illness/ Impairment	A09	Output Problem		
							F1905	Device Revision or Replacement	A09	Output Problem		
					CT2750	Implantable cardiac electrical stimulators and associated devices	E0112	Dizziness	A09	Output Problem		
							E2403	No Clinical Signs, Symptoms or Conditions	A09	Output Problem		
							F08	Hospitalization or Prolonged Hospitalization	A09	Output Problem		
							F11	Non-Serious Injury/ Illness/ Impairment	A09	Output Problem		
							F1905	Device Revision or Replacement	A09	Output Problem		
					CT2866	Implantable cardiac electrical stimulators	E0112	Dizziness	A09	Output Problem		
							E2403	No Clinical Signs, Symptoms or Conditions	A09	Output Problem		
							F08	Hospitalization or Prolonged Hospitalization	A09	Output Problem		
							F11	Non-Serious Injury/ Illness/ Impairment	A09	Output Problem		
							F1905	Device Revision or Replacement	A09	Output Problem		
					CT2985	Active	E0112	Dizziness	A09	Output Problem		
							E2403	No Clinical Signs, Symptoms or Conditions	A09	Output Problem		
							F08	Hospitalization or Prolonged Hospitalization	A09	Output Problem		
							F11	Non-Serious Injury/ Illness/ Impairment	A09	Output Problem		
							F1905	Device Revision or Replacement	A09	Output Problem		
					CT3001	Implantable metal	E0112	Dizziness	A09	Output Problem		
							E2403	No Clinical Signs, Symptoms or Conditions	A09	Output Problem		
							F08	Hospitalization or Prolonged Hospitalization	A09	Output Problem		
							F11	Non-Serious Injury/ Illness/ Impairment	A09	Output Problem		
							F1905	Device Revision or Replacement	A09	Output Problem		
					CT3002	Implantable synthetic polymer	E0112	Dizziness	A09	Output Problem		
							E2403	No Clinical Signs, Symptoms or Conditions	A09	Output Problem		
							F08	Hospitalization or Prolonged Hospitalization	A09	Output Problem		
							F11	Non-Serious Injury/ Illness/ Impairment	A09	Output Problem		
							F1905	Device Revision or Replacement	A09	Output Problem		
					CT3006	Implantable inorganic material	E0112	Dizziness	A09	Output Problem		
							E2403	No Clinical Signs, Symptoms or Conditions	A09	Output Problem		
							F08	Hospitalization or Prolonged Hospitalization	A09	Output Problem		
							F11	Non-Serious Injury/ Illness/ Impairment	A09	Output Problem		
							F1905	Device Revision or Replacement	A09	Output Problem		
					CT314	Electrically powered	E0112	Dizziness	A09	Output Problem		
							E2403	No Clinical Signs, Symptoms or Conditions	A09	Output Problem		
							F08	Hospitalization or Prolonged Hospitalization	A09	Output Problem		
							F11	Non-Serious Injury/ Illness/ Impairment	A09	Output Problem		
							F1905	Device Revision or Replacement	A09	Output Problem		
					CT321	Long-term surgical invasive	E0112	Dizziness	A09	Output Problem		
							E2403	No Clinical Signs, Symptoms or Conditions	A09	Output Problem		
							F08	Hospitalization or Prolonged Hospitalization	A09	Output Problem		
							F11	Non-Serious Injury/ Illness/ Impairment	A09	Output Problem		
							F1905	Device Revision or Replacement	A09	Output Problem		
					CT449	Generators and associated devices	E0112	Dizziness	A09	Output Problem		
							E2403	No Clinical Signs, Symptoms or Conditions	A09	Output Problem		
							F08	Hospitalization or Prolonged Hospitalization	A09	Output Problem		
							F11	Non-Serious Injury/ Illness/ Impairment	A09	Output Problem		
							F1905	Device Revision or Replacement	A09	Output Problem		
					CT981	Single use	E0112	Dizziness	A09	Output Problem		
							E2403	No Clinical Signs, Symptoms or Conditions	A09	Output Problem		
							F08	Hospitalization or Prolonged Hospitalization	A09	Output Problem		
							F11	Non-Serious Injury/ Illness/ Impairment	A09	Output Problem		
							F1905	Device Revision or Replacement	A09	Output Problem		
					CT983	Surgical invasive	E0112	Dizziness	A09	Output Problem		
							E2403	No Clinical Signs, Symptoms or Conditions	A09	Output Problem		
							F08	Hospitalization or Prolonged Hospitalization	A09	Output Problem		
							F11	Non-Serious Injury/ Illness/ Impairment	A09	Output Problem		

							F1905	Device Revision or Replacement	A09	Output Problem			
6365	2022	All other reportable incidents	35039	Epicardial pacing lead	CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070909	Oversensing			
									A090801	Signal Artifact/Noise			
							F26	No Health Consequences or Impact	A070909	Oversensing			
									A090801	Signal Artifact/Noise			
6366	2022	All other reportable incidents	47265	Dual-chamber implantable pacemaker, rate-responsive	CT141	Electrodes/Leads and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	15/10/2021	15/10/2021	
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	15/10/2021	15/10/2021	
							F1903	Device Explantation	A070504	Premature Discharge of Battery	15/10/2021	15/10/2021	
					CT145	Cardiovascular devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	15/10/2021	15/10/2021	
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	15/10/2021	15/10/2021	
							F1903	Device Explantation	A070504	Premature Discharge of Battery	15/10/2021	15/10/2021	
					CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	15/10/2021	15/10/2021	
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	15/10/2021	15/10/2021	
							F1903	Device Explantation	A070504	Premature Discharge of Battery	15/10/2021	15/10/2021	
					CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	15/10/2021	15/10/2021	
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	15/10/2021	15/10/2021	
							F1903	Device Explantation	A070504	Premature Discharge of Battery	15/10/2021	15/10/2021	
					CT233	Surgical	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	15/10/2021	15/10/2021	
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	15/10/2021	15/10/2021	
							F1903	Device Explantation	A070504	Premature Discharge of Battery	15/10/2021	15/10/2021	
					CT2406	Implantable	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	15/10/2021	15/10/2021	
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	15/10/2021	15/10/2021	
							F1903	Device Explantation	A070504	Premature Discharge of Battery	15/10/2021	15/10/2021	
					CT2407	Active implantable	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	15/10/2021	15/10/2021	
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	15/10/2021	15/10/2021	
							F1903	Device Explantation	A070504	Premature Discharge of Battery	15/10/2021	15/10/2021	
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	15/10/2021	15/10/2021	
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	15/10/2021	15/10/2021	
							F1903	Device Explantation	A070504	Premature Discharge of Battery	15/10/2021	15/10/2021	
					CT2727	Active implant devices/ancillaries	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	15/10/2021	15/10/2021	
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	15/10/2021	15/10/2021	
							F1903	Device Explantation	A070504	Premature Discharge of Battery	15/10/2021	15/10/2021	
					CT2750	Implantable cardiac electrical stimulators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	15/10/2021	15/10/2021	
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	15/10/2021	15/10/2021	
							F1903	Device Explantation	A070504	Premature Discharge of Battery	15/10/2021	15/10/2021	
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	15/10/2021	15/10/2021	
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	15/10/2021	15/10/2021	
							F1903	Device Explantation	A070504	Premature Discharge of Battery	15/10/2021	15/10/2021	
					CT2985	Active	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	15/10/2021	15/10/2021	
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	15/10/2021	15/10/2021	
							F1903	Device Explantation	A070504	Premature Discharge of Battery	15/10/2021	15/10/2021	
					CT3001	Implantable metal	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	15/10/2021	15/10/2021	
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	15/10/2021	15/10/2021	
							F1903	Device Explantation	A070504	Premature Discharge of Battery	15/10/2021	15/10/2021	
					CT3002	Implantable synthetic polymer	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	15/10/2021	15/10/2021	
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	15/10/2021	15/10/2021	
							F1903	Device Explantation	A070504	Premature Discharge of Battery	15/10/2021	15/10/2021	
					CT3006	Implantable inorganic material	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	15/10/2021	15/10/2021	
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	15/10/2021	15/10/2021	
							F1903	Device Explantation	A070504	Premature Discharge of Battery	15/10/2021	15/10/2021	
					CT314	Electrically powered	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	15/10/2021	15/10/2021	
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	15/10/2021	15/10/2021	
							F1903	Device Explantation	A070504	Premature Discharge of Battery	15/10/2021	15/10/2021	
					CT321	Long-term surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	15/10/2021	15/10/2021	
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	15/10/2021	15/10/2021	
							F1903	Device Explantation	A070504	Premature Discharge of Battery	15/10/2021	15/10/2021	
					CT449	Generators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	15/10/2021	15/10/2021	
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	15/10/2021	15/10/2021	
							F1903	Device Explantation	A070504	Premature Discharge of Battery	15/10/2021	15/10/2021	
					CT981	Single use	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	15/10/2021	15/10/2021	
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	15/10/2021	15/10/2021	
							F1903	Device Explantation	A070504	Premature Discharge of Battery	15/10/2021	15/10/2021	
					CT983	Surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	15/10/2021	15/10/2021	
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	15/10/2021	15/10/2021	
							F1903	Device Explantation	A070504	Premature Discharge of Battery	15/10/2021	15/10/2021	
6367	2022	All other reportable incidents	35852	Single-chamber implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem	13/09/2022	13/09/2022	
									A070504	Premature Discharge of Battery	13/09/2022	13/09/2022	
							F08	Hospitalization or Prolonged Hospitalization	A0705	Battery Problem	13/09/2022	13/09/2022	
									A070504	Premature Discharge of Battery	13/09/2022	13/09/2022	
							F1905	Device Revision or Replacement	A0705	Battery Problem	13/09/2022	13/09/2022	
									A070504	Premature Discharge of Battery	13/09/2022	13/09/2022	
6368	2021	All other reportable incidents	47265	Dual-chamber implantable pacemaker, rate-responsive	CT141	Electrodes/Leads and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070909	Oversensing	13/07/2021	13/07/2021	
							F1905	Device Revision or Replacement	A070909	Oversensing	13/07/2021	13/07/2021	
					CT145	Cardiovascular devices	E2403	No Clinical Signs, Symptoms or Conditions	A070909	Oversensing	13/07/2021	13/07/2021	
							F1905	Device Revision or Replacement	A070909	Oversensing	13/07/2021	13/07/2021	
					CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070909	Oversensing	13/07/2021	13/07/2021	
							F1905	Device Revision or Replacement	A070909	Oversensing	13/07/2021	13/07/2021	
					CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A070909	Oversensing	13/07/2021	13/07/2021	
							F1905	Device Revision or Replacement	A070909	Oversensing	13/07/2021	13/07/2021	
					CT233	Surgical	E2403	No Clinical Signs, Symptoms or Conditions	A070909	Oversensing	13/07/2021	13/07/2021	
							F1905	Device Revision or Replacement	A070909	Oversensing	13/07/2021	13/07/2021	
					CT2406	Implantable	E2403	No Clinical Signs, Symptoms or Conditions	A070909	Oversensing	13/07/2021	13/07/2021	
							F1905	Device Revision or Replacement	A070909	Oversensing	13/07/2021	13/07/2021	
					CT2407	Active implantable	E2403	No Clinical Signs, Symptoms or Conditions	A070909	Oversensing	13/07/2021	13/07/2021	
							F1905	Device Revision or Replacement	A070909	Oversensing	13/07/2021	13/07/2021	
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A070909	Oversensing	13/07/2021	13/07/2021	
							F1905	Device Revision or Replacement	A070909	Oversensing	13/07/2021	13/07/2021	
					CT2727	Active implant devices/ancillaries	E2403	No Clinical Signs, Symptoms or Conditions	A070909	Oversensing	13/07/2021	13/07/2021	

						F1905	Device Revision or Replacement	A070909	Oversensing	13/07/2021	13/07/2021
				CT2750	Implantable cardiac electrical stimulators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070909	Oversensing	13/07/2021	13/07/2021
						F1905	Device Revision or Replacement	A070909	Oversensing	13/07/2021	13/07/2021
				CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A070909	Oversensing	13/07/2021	13/07/2021
						F1905	Device Revision or Replacement	A070909	Oversensing	13/07/2021	13/07/2021
				CT2985	Active	E2403	No Clinical Signs, Symptoms or Conditions	A070909	Oversensing	13/07/2021	13/07/2021
						F1905	Device Revision or Replacement	A070909	Oversensing	13/07/2021	13/07/2021
				CT3001	Implantable metal	E2403	No Clinical Signs, Symptoms or Conditions	A070909	Oversensing	13/07/2021	13/07/2021
						F1905	Device Revision or Replacement	A070909	Oversensing	13/07/2021	13/07/2021
				CT3002	Implantable synthetic polymer	E2403	No Clinical Signs, Symptoms or Conditions	A070909	Oversensing	13/07/2021	13/07/2021
						F1905	Device Revision or Replacement	A070909	Oversensing	13/07/2021	13/07/2021
				CT3006	Implantable inorganic material	E2403	No Clinical Signs, Symptoms or Conditions	A070909	Oversensing	13/07/2021	13/07/2021
						F1905	Device Revision or Replacement	A070909	Oversensing	13/07/2021	13/07/2021
				CT314	Electrically powered	E2403	No Clinical Signs, Symptoms or Conditions	A070909	Oversensing	13/07/2021	13/07/2021
						F1905	Device Revision or Replacement	A070909	Oversensing	13/07/2021	13/07/2021
				CT321	Long-term surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A070909	Oversensing	13/07/2021	13/07/2021
						F1905	Device Revision or Replacement	A070909	Oversensing	13/07/2021	13/07/2021
				CT449	Generators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070909	Oversensing	13/07/2021	13/07/2021
						F1905	Device Revision or Replacement	A070909	Oversensing	13/07/2021	13/07/2021
				CT981	Single use	E2403	No Clinical Signs, Symptoms or Conditions	A070909	Oversensing	13/07/2021	13/07/2021
						F1905	Device Revision or Replacement	A070909	Oversensing	13/07/2021	13/07/2021
				CT983	Surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A070909	Oversensing	13/07/2021	13/07/2021
						F1905	Device Revision or Replacement	A070909	Oversensing	13/07/2021	13/07/2021
6369	2021	All other reportable incidents	17882	Physiologic-monitoring defibrillation system	CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A07	Electrical /Electronic Property Problem		
							F27	Problem Identified During Non-Clinical Procedure	A07	Electrical /Electronic Property Problem		
6370	2021	All other reportable incidents	36241	Pacemaker/defibrillator lead	CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A040101	Fracture		
									A070909	Oversensing		
									A0712	Pacing Problem		
									A072201	High Impedance		
							F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A040101	Fracture		
									A070909	Oversensing		
									A0712	Pacing Problem		
									A072201	High Impedance		
							F23	Unexpected Medical Intervention	A040101	Fracture		
									A070909	Oversensing		
									A0712	Pacing Problem		
									A072201	High Impedance		
6371	2021	All other reportable incidents	47205	Cardiac pulse generator programmer	CT1684	Cardiac pacemakers and associated devices	E060109	Tachycardia	A0902	Display or Visual Feedback Problem		
							E2403	No Clinical Signs, Symptoms or Conditions	A0902	Display or Visual Feedback Problem		
							F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A0902	Display or Visual Feedback Problem		
6372	2021	All other reportable incidents	47265	Dual-chamber implantable pacemaker, rate-responsive	CT141	Electrodes/Leads and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	27/10/2021	27/10/2021
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	27/10/2021	27/10/2021
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	27/10/2021	27/10/2021
					CT145	Cardiovascular devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	27/10/2021	27/10/2021
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	27/10/2021	27/10/2021
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	27/10/2021	27/10/2021
					CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	27/10/2021	27/10/2021
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	27/10/2021	27/10/2021
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	27/10/2021	27/10/2021
					CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	27/10/2021	27/10/2021
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	27/10/2021	27/10/2021
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	27/10/2021	27/10/2021
					CT233	Surgical	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	27/10/2021	27/10/2021
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	27/10/2021	27/10/2021
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	27/10/2021	27/10/2021
					CT2406	Implantable	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	27/10/2021	27/10/2021
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	27/10/2021	27/10/2021
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	27/10/2021	27/10/2021
					CT2407	Active implantable	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	27/10/2021	27/10/2021
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	27/10/2021	27/10/2021
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	27/10/2021	27/10/2021
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	27/10/2021	27/10/2021
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	27/10/2021	27/10/2021
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	27/10/2021	27/10/2021
					CT2727	Active implant devices/ancillaries	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	27/10/2021	27/10/2021
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	27/10/2021	27/10/2021
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	27/10/2021	27/10/2021
					CT2750	Implantable cardiac electrical stimulators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	27/10/2021	27/10/2021
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	27/10/2021	27/10/2021
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	27/10/2021	27/10/2021
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	27/10/2021	27/10/2021
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	27/10/2021	27/10/2021
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	27/10/2021	27/10/2021
					CT2985	Active	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	27/10/2021	27/10/2021
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	27/10/2021	27/10/2021
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	27/10/2021	27/10/2021
					CT3001	Implantable metal	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	27/10/2021	27/10/2021
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	27/10/2021	27/10/2021
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	27/10/2021	27/10/2021
					CT3002	Implantable synthetic polymer	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	27/10/2021	27/10/2021
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	27/10/2021	27/10/2021
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	27/10/2021	27/10/2021
					CT3006	Implantable inorganic material	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	27/10/2021	27/10/2021
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	27/10/2021	27/10/2021
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	27/10/2021	27/10/2021
					CT314	Electrically powered	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	27/10/2021	27/10/2021
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	27/10/2021	27/10/2021
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	27/10/2021	27/10/2021

					CT321	Long-term surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	27/10/2021	27/10/2021
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	27/10/2021	27/10/2021
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	27/10/2021	27/10/2021
					CT449	Generators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	27/10/2021	27/10/2021
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	27/10/2021	27/10/2021
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	27/10/2021	27/10/2021
					CT981	Single use	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	27/10/2021	27/10/2021
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	27/10/2021	27/10/2021
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	27/10/2021	27/10/2021
					CT983	Surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	27/10/2021	27/10/2021
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	27/10/2021	27/10/2021
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	27/10/2021	27/10/2021
6373	2021	All other reportable incidents	35852	Single-chamber implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem	18/08/2021	18/08/2021
									A070504	Premature Discharge of Battery	18/08/2021	18/08/2021
							F08	Hospitalization or Prolonged Hospitalization	A0705	Battery Problem	18/08/2021	18/08/2021
									A070504	Premature Discharge of Battery	18/08/2021	18/08/2021
							F1905	Device Revision or Replacement	A0705	Battery Problem	18/08/2021	18/08/2021
									A070504	Premature Discharge of Battery	18/08/2021	18/08/2021
6374	2021	All other reportable incidents	35852	Single-chamber implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem	27/08/2021	27/08/2021
									A070504	Premature Discharge of Battery	27/08/2021	27/08/2021
							F08	Hospitalization or Prolonged Hospitalization	A0705	Battery Problem	27/08/2021	27/08/2021
									A070504	Premature Discharge of Battery	27/08/2021	27/08/2021
							F1903	Device Explantation	A0705	Battery Problem	27/08/2021	27/08/2021
									A070504	Premature Discharge of Battery	27/08/2021	27/08/2021
6375	2021	All other reportable incidents	35852	Single-chamber implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem	21/09/2021	21/09/2021
									A070504	Premature Discharge of Battery	21/09/2021	21/09/2021
							F08	Hospitalization or Prolonged Hospitalization	A0705	Battery Problem	21/09/2021	21/09/2021
									A070504	Premature Discharge of Battery	21/09/2021	21/09/2021
							F1905	Device Revision or Replacement	A0705	Battery Problem	21/09/2021	21/09/2021
									A070504	Premature Discharge of Battery	21/09/2021	21/09/2021
6376	2021	All other reportable incidents	37265	Dual-chamber implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	23/09/2021	23/09/2021
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	23/09/2021	23/09/2021
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	23/09/2021	23/09/2021
6377	2021	All other reportable incidents	47265	Dual-chamber implantable pacemaker, rate-responsive	CT141	Electrodes/Leads and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A0711	Interrogation Problem		
									A1304	Telemetry Discrepancy		
							F26	No Health Consequences or Impact	A0711	Interrogation Problem		
									A1304	Telemetry Discrepancy		
					CT145	Cardiovascular devices	E2403	No Clinical Signs, Symptoms or Conditions	A0711	Interrogation Problem		
									A1304	Telemetry Discrepancy		
							F26	No Health Consequences or Impact	A0711	Interrogation Problem		
									A1304	Telemetry Discrepancy		
					CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A0711	Interrogation Problem		
									A1304	Telemetry Discrepancy		
							F26	No Health Consequences or Impact	A0711	Interrogation Problem		
									A1304	Telemetry Discrepancy		
					CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A0711	Interrogation Problem		
									A1304	Telemetry Discrepancy		
							F26	No Health Consequences or Impact	A0711	Interrogation Problem		
									A1304	Telemetry Discrepancy		
					CT233	Surgical	E2403	No Clinical Signs, Symptoms or Conditions	A0711	Interrogation Problem		
									A1304	Telemetry Discrepancy		
							F26	No Health Consequences or Impact	A0711	Interrogation Problem		
									A1304	Telemetry Discrepancy		
					CT2406	Implantable	E2403	No Clinical Signs, Symptoms or Conditions	A0711	Interrogation Problem		
									A1304	Telemetry Discrepancy		
							F26	No Health Consequences or Impact	A0711	Interrogation Problem		
									A1304	Telemetry Discrepancy		
					CT2407	Active implantable	E2403	No Clinical Signs, Symptoms or Conditions	A0711	Interrogation Problem		
									A1304	Telemetry Discrepancy		
							F26	No Health Consequences or Impact	A0711	Interrogation Problem		
									A1304	Telemetry Discrepancy		
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A0711	Interrogation Problem		
									A1304	Telemetry Discrepancy		
							F26	No Health Consequences or Impact	A0711	Interrogation Problem		
									A1304	Telemetry Discrepancy		
					CT2727	Active implant devices/ancillaries	E2403	No Clinical Signs, Symptoms or Conditions	A0711	Interrogation Problem		
									A1304	Telemetry Discrepancy		
							F26	No Health Consequences or Impact	A0711	Interrogation Problem		
									A1304	Telemetry Discrepancy		
					CT2750	Implantable cardiac electrical stimulators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A0711	Interrogation Problem		
									A1304	Telemetry Discrepancy		
							F26	No Health Consequences or Impact	A0711	Interrogation Problem		
									A1304	Telemetry Discrepancy		
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A0711	Interrogation Problem		
									A1304	Telemetry Discrepancy		
							F26	No Health Consequences or Impact	A0711	Interrogation Problem		
									A1304	Telemetry Discrepancy		
					CT2985	Active	E2403	No Clinical Signs, Symptoms or Conditions	A0711	Interrogation Problem		
									A1304	Telemetry Discrepancy		
							F26	No Health Consequences or Impact	A0711	Interrogation Problem		
									A1304	Telemetry Discrepancy		
					CT3001	Implantable metal	E2403	No Clinical Signs, Symptoms or Conditions	A0711	Interrogation Problem		
									A1304	Telemetry Discrepancy		
							F26	No Health Consequences or Impact	A0711	Interrogation Problem		
									A1304	Telemetry Discrepancy		
					CT3002	Implantable synthetic polymer	E2403	No Clinical Signs, Symptoms or Conditions	A0711	Interrogation Problem		
									A1304	Telemetry Discrepancy		
							F26	No Health Consequences or Impact	A0711	Interrogation Problem		

									A1304	Telemetry Discrepancy		
					CT3006	Implantable inorganic material	E2403	No Clinical Signs, Symptoms or Conditions	A0711	Interrogation Problem		
									A1304	Telemetry Discrepancy		
							F26	No Health Consequences or Impact	A0711	Interrogation Problem		
									A1304	Telemetry Discrepancy		
					CT314	Electrically powered	E2403	No Clinical Signs, Symptoms or Conditions	A0711	Interrogation Problem		
									A1304	Telemetry Discrepancy		
							F26	No Health Consequences or Impact	A0711	Interrogation Problem		
									A1304	Telemetry Discrepancy		
					CT321	Long-term surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A0711	Interrogation Problem		
									A1304	Telemetry Discrepancy		
							F26	No Health Consequences or Impact	A0711	Interrogation Problem		
									A1304	Telemetry Discrepancy		
					CT449	Generators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A0711	Interrogation Problem		
									A1304	Telemetry Discrepancy		
							F26	No Health Consequences or Impact	A0711	Interrogation Problem		
									A1304	Telemetry Discrepancy		
					CT981	Single use	E2403	No Clinical Signs, Symptoms or Conditions	A0711	Interrogation Problem		
									A1304	Telemetry Discrepancy		
							F26	No Health Consequences or Impact	A0711	Interrogation Problem		
									A1304	Telemetry Discrepancy		
					CT983	Surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A0711	Interrogation Problem		
									A1304	Telemetry Discrepancy		
							F26	No Health Consequences or Impact	A0711	Interrogation Problem		
									A1304	Telemetry Discrepancy		
6378	2022	All other reportable incidents	37265	Dual-chamber implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	16/03/2022	16/03/2022
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	16/03/2022	16/03/2022
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	16/03/2022	16/03/2022
6379	2022	All other reportable incidents	47265	Dual-chamber implantable pacemaker, rate-responsive	CT141	Electrodes/Leads and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
							F24	Insufficient Health Impact Information	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
					CT145	Cardiovascular devices	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
							F24	Insufficient Health Impact Information	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
					CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
							F24	Insufficient Health Impact Information	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
					CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
							F24	Insufficient Health Impact Information	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
					CT233	Surgical	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
							F24	Insufficient Health Impact Information	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
					CT2406	Implantable	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
							F24	Insufficient Health Impact Information	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
					CT2407	Active implantable	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
							F24	Insufficient Health Impact Information	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
							F24	Insufficient Health Impact Information	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
					CT2727	Active implant devices/ancillaries	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
							F24	Insufficient Health Impact Information	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
					CT2750	Implantable cardiac electrical stimulators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
							F24	Insufficient Health Impact Information	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
							F24	Insufficient Health Impact Information	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
					CT2985	Active	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
							F24	Insufficient Health Impact Information	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
					CT3001	Implantable metal	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
							F24	Insufficient Health Impact Information	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
					CT3002	Implantable synthetic polymer	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
							F24	Insufficient Health Impact Information	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
					CT3006	Implantable inorganic material	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
							F24	Insufficient Health Impact Information	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		

					CT314	Electrically powered	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
							F24	Insufficient Health Impact Information	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
					CT321	Long-term surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
							F24	Insufficient Health Impact Information	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
					CT449	Generators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
							F24	Insufficient Health Impact Information	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
					CT981	Single use	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
							F24	Insufficient Health Impact Information	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
					CT983	Surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
							F24	Insufficient Health Impact Information	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
6380	2021	All other reportable incidents	47270	Cardiac resynchronization therapy implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
									A1605	Reset Problem		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
									A1605	Reset Problem		
6381	2022	All other reportable incidents	36241	Pacemaker/defibrillator lead	CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070909	Oversensing		
									A072201	High Impedance		
									A090801	Signal Artifact/Noise		
							F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A070909	Oversensing		
									A072201	High Impedance		
									A090801	Signal Artifact/Noise		
6382	2022	All other reportable incidents	47265	Dual-chamber implantable pacemaker, rate-responsive	CT141	Electrodes/Leads and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A0203	Defective Device	13/04/2022	13/04/2022
									A0705	Battery Problem	13/04/2022	13/04/2022
									A070504	Premature Discharge of Battery	13/04/2022	13/04/2022
							F08	Hospitalization or Prolonged Hospitalization	A0203	Defective Device	13/04/2022	13/04/2022
									A0705	Battery Problem	13/04/2022	13/04/2022
									A070504	Premature Discharge of Battery	13/04/2022	13/04/2022
							F1905	Device Revision or Replacement	A0203	Defective Device	13/04/2022	13/04/2022
									A0705	Battery Problem	13/04/2022	13/04/2022
									A070504	Premature Discharge of Battery	13/04/2022	13/04/2022
					CT145	Cardiovascular devices	E2403	No Clinical Signs, Symptoms or Conditions	A0203	Defective Device	13/04/2022	13/04/2022
									A0705	Battery Problem	13/04/2022	13/04/2022
									A070504	Premature Discharge of Battery	13/04/2022	13/04/2022
							F08	Hospitalization or Prolonged Hospitalization	A0203	Defective Device	13/04/2022	13/04/2022
									A0705	Battery Problem	13/04/2022	13/04/2022
									A070504	Premature Discharge of Battery	13/04/2022	13/04/2022
							F1905	Device Revision or Replacement	A0203	Defective Device	13/04/2022	13/04/2022
									A0705	Battery Problem	13/04/2022	13/04/2022
									A070504	Premature Discharge of Battery	13/04/2022	13/04/2022
					CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A0203	Defective Device	13/04/2022	13/04/2022
									A0705	Battery Problem	13/04/2022	13/04/2022
									A070504	Premature Discharge of Battery	13/04/2022	13/04/2022
							F08	Hospitalization or Prolonged Hospitalization	A0203	Defective Device	13/04/2022	13/04/2022
									A0705	Battery Problem	13/04/2022	13/04/2022
									A070504	Premature Discharge of Battery	13/04/2022	13/04/2022
							F1905	Device Revision or Replacement	A0203	Defective Device	13/04/2022	13/04/2022
									A0705	Battery Problem	13/04/2022	13/04/2022
									A070504	Premature Discharge of Battery	13/04/2022	13/04/2022
					CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A0203	Defective Device	13/04/2022	13/04/2022
									A0705	Battery Problem	13/04/2022	13/04/2022
									A070504	Premature Discharge of Battery	13/04/2022	13/04/2022
							F08	Hospitalization or Prolonged Hospitalization	A0203	Defective Device	13/04/2022	13/04/2022
									A0705	Battery Problem	13/04/2022	13/04/2022
									A070504	Premature Discharge of Battery	13/04/2022	13/04/2022
							F1905	Device Revision or Replacement	A0203	Defective Device	13/04/2022	13/04/2022
									A0705	Battery Problem	13/04/2022	13/04/2022
									A070504	Premature Discharge of Battery	13/04/2022	13/04/2022
					CT233	Surgical	E2403	No Clinical Signs, Symptoms or Conditions	A0203	Defective Device	13/04/2022	13/04/2022
									A0705	Battery Problem	13/04/2022	13/04/2022
									A070504	Premature Discharge of Battery	13/04/2022	13/04/2022
							F08	Hospitalization or Prolonged Hospitalization	A0203	Defective Device	13/04/2022	13/04/2022
									A0705	Battery Problem	13/04/2022	13/04/2022
									A070504	Premature Discharge of Battery	13/04/2022	13/04/2022
							F1905	Device Revision or Replacement	A0203	Defective Device	13/04/2022	13/04/2022
									A0705	Battery Problem	13/04/2022	13/04/2022
									A070504	Premature Discharge of Battery	13/04/2022	13/04/2022
					CT2406	Implantable	E2403	No Clinical Signs, Symptoms or Conditions	A0203	Defective Device	13/04/2022	13/04/2022
									A0705	Battery Problem	13/04/2022	13/04/2022
									A070504	Premature Discharge of Battery	13/04/2022	13/04/2022
							F08	Hospitalization or Prolonged Hospitalization	A0203	Defective Device	13/04/2022	13/04/2022
									A0705	Battery Problem	13/04/2022	13/04/2022
									A070504	Premature Discharge of Battery	13/04/2022	13/04/2022
							F1905	Device Revision or Replacement	A0203	Defective Device	13/04/2022	13/04/2022
									A0705	Battery Problem	13/04/2022	13/04/2022
									A070504	Premature Discharge of Battery	13/04/2022	13/04/2022
					CT2407	Active implantable	E2403	No Clinical Signs, Symptoms or Conditions	A0203	Defective Device	13/04/2022	13/04/2022
									A0705	Battery Problem	13/04/2022	13/04/2022
									A070504	Premature Discharge of Battery	13/04/2022	13/04/2022
							F08	Hospitalization or Prolonged Hospitalization	A0203	Defective Device	13/04/2022	13/04/2022

								A0705	Battery Problem	13/04/2022	13/04/2022
								A070504	Premature Discharge of Battery	13/04/2022	13/04/2022
						F1905	Device Revision or Replacement	A0203	Defective Device	13/04/2022	13/04/2022
								A0705	Battery Problem	13/04/2022	13/04/2022
								A070504	Premature Discharge of Battery	13/04/2022	13/04/2022
					CT257	Implantable cardiac pacemakers	E2403	A0203	Defective Device	13/04/2022	13/04/2022
								A0705	Battery Problem	13/04/2022	13/04/2022
								A070504	Premature Discharge of Battery	13/04/2022	13/04/2022
							F08	A0203	Defective Device	13/04/2022	13/04/2022
								A0705	Battery Problem	13/04/2022	13/04/2022
								A070504	Premature Discharge of Battery	13/04/2022	13/04/2022
							F1905	A0203	Defective Device	13/04/2022	13/04/2022
								A0705	Battery Problem	13/04/2022	13/04/2022
								A070504	Premature Discharge of Battery	13/04/2022	13/04/2022
					CT2727	Active implant devices/ancillaries	E2403	A0203	Defective Device	13/04/2022	13/04/2022
								A0705	Battery Problem	13/04/2022	13/04/2022
								A070504	Premature Discharge of Battery	13/04/2022	13/04/2022
							F08	A0203	Defective Device	13/04/2022	13/04/2022
								A0705	Battery Problem	13/04/2022	13/04/2022
								A070504	Premature Discharge of Battery	13/04/2022	13/04/2022
							F1905	A0203	Defective Device	13/04/2022	13/04/2022
								A0705	Battery Problem	13/04/2022	13/04/2022
								A070504	Premature Discharge of Battery	13/04/2022	13/04/2022
					CT2750	Implantable cardiac electrical stimulators and associated devices	E2403	A0203	Defective Device	13/04/2022	13/04/2022
								A0705	Battery Problem	13/04/2022	13/04/2022
								A070504	Premature Discharge of Battery	13/04/2022	13/04/2022
							F08	A0203	Defective Device	13/04/2022	13/04/2022
								A0705	Battery Problem	13/04/2022	13/04/2022
								A070504	Premature Discharge of Battery	13/04/2022	13/04/2022
							F1905	A0203	Defective Device	13/04/2022	13/04/2022
								A0705	Battery Problem	13/04/2022	13/04/2022
								A070504	Premature Discharge of Battery	13/04/2022	13/04/2022
					CT2866	Implantable cardiac electrical stimulators	E2403	A0203	Defective Device	13/04/2022	13/04/2022
								A0705	Battery Problem	13/04/2022	13/04/2022
								A070504	Premature Discharge of Battery	13/04/2022	13/04/2022
							F08	A0203	Defective Device	13/04/2022	13/04/2022
								A0705	Battery Problem	13/04/2022	13/04/2022
								A070504	Premature Discharge of Battery	13/04/2022	13/04/2022
							F1905	A0203	Defective Device	13/04/2022	13/04/2022
								A0705	Battery Problem	13/04/2022	13/04/2022
								A070504	Premature Discharge of Battery	13/04/2022	13/04/2022
					CT2985	Active	E2403	A0203	Defective Device	13/04/2022	13/04/2022
								A0705	Battery Problem	13/04/2022	13/04/2022
								A070504	Premature Discharge of Battery	13/04/2022	13/04/2022
							F08	A0203	Defective Device	13/04/2022	13/04/2022
								A0705	Battery Problem	13/04/2022	13/04/2022
								A070504	Premature Discharge of Battery	13/04/2022	13/04/2022
							F1905	A0203	Defective Device	13/04/2022	13/04/2022
								A0705	Battery Problem	13/04/2022	13/04/2022
								A070504	Premature Discharge of Battery	13/04/2022	13/04/2022
					CT3001	Implantable metal	E2403	A0203	Defective Device	13/04/2022	13/04/2022
								A0705	Battery Problem	13/04/2022	13/04/2022
								A070504	Premature Discharge of Battery	13/04/2022	13/04/2022
							F08	A0203	Defective Device	13/04/2022	13/04/2022
								A0705	Battery Problem	13/04/2022	13/04/2022
								A070504	Premature Discharge of Battery	13/04/2022	13/04/2022
							F1905	A0203	Defective Device	13/04/2022	13/04/2022
								A0705	Battery Problem	13/04/2022	13/04/2022
								A070504	Premature Discharge of Battery	13/04/2022	13/04/2022
					CT3002	Implantable synthetic polymer	E2403	A0203	Defective Device	13/04/2022	13/04/2022
								A0705	Battery Problem	13/04/2022	13/04/2022
								A070504	Premature Discharge of Battery	13/04/2022	13/04/2022
							F08	A0203	Defective Device	13/04/2022	13/04/2022
								A0705	Battery Problem	13/04/2022	13/04/2022
								A070504	Premature Discharge of Battery	13/04/2022	13/04/2022
							F1905	A0203	Defective Device	13/04/2022	13/04/2022
								A0705	Battery Problem	13/04/2022	13/04/2022
								A070504	Premature Discharge of Battery	13/04/2022	13/04/2022
					CT3006	Implantable inorganic material	E2403	A0203	Defective Device	13/04/2022	13/04/2022
								A0705	Battery Problem	13/04/2022	13/04/2022
								A070504	Premature Discharge of Battery	13/04/2022	13/04/2022
							F08	A0203	Defective Device	13/04/2022	13/04/2022
								A0705	Battery Problem	13/04/2022	13/04/2022
								A070504	Premature Discharge of Battery	13/04/2022	13/04/2022
							F1905	A0203	Defective Device	13/04/2022	13/04/2022
								A0705	Battery Problem	13/04/2022	13/04/2022
								A070504	Premature Discharge of Battery	13/04/2022	13/04/2022
					CT314	Electrically powered	E2403	A0203	Defective Device	13/04/2022	13/04/2022
								A0705	Battery Problem	13/04/2022	13/04/2022
								A070504	Premature Discharge of Battery	13/04/2022	13/04/2022
							F08	A0203	Defective Device	13/04/2022	13/04/2022
								A0705	Battery Problem	13/04/2022	13/04/2022
								A070504	Premature Discharge of Battery	13/04/2022	13/04/2022
							F1905	A0203	Defective Device	13/04/2022	13/04/2022
								A0705	Battery Problem	13/04/2022	13/04/2022
								A070504	Premature Discharge of Battery	13/04/2022	13/04/2022
					CT321	Long-term surgical invasive	E2403	A0203	Defective Device	13/04/2022	13/04/2022
								A0705	Battery Problem	13/04/2022	13/04/2022

									A070504	Premature Discharge of Battery	13/04/2022	13/04/2022
							F08	Hospitalization or Prolonged Hospitalization	A0203	Defective Device	13/04/2022	13/04/2022
									A0705	Battery Problem	13/04/2022	13/04/2022
									A070504	Premature Discharge of Battery	13/04/2022	13/04/2022
							F1905	Device Revision or Replacement	A0203	Defective Device	13/04/2022	13/04/2022
									A0705	Battery Problem	13/04/2022	13/04/2022
									A070504	Premature Discharge of Battery	13/04/2022	13/04/2022
					CT449	Generators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A0203	Defective Device	13/04/2022	13/04/2022
									A0705	Battery Problem	13/04/2022	13/04/2022
									A070504	Premature Discharge of Battery	13/04/2022	13/04/2022
							F08	Hospitalization or Prolonged Hospitalization	A0203	Defective Device	13/04/2022	13/04/2022
									A0705	Battery Problem	13/04/2022	13/04/2022
									A070504	Premature Discharge of Battery	13/04/2022	13/04/2022
							F1905	Device Revision or Replacement	A0203	Defective Device	13/04/2022	13/04/2022
									A0705	Battery Problem	13/04/2022	13/04/2022
									A070504	Premature Discharge of Battery	13/04/2022	13/04/2022
					CT981	Single use	E2403	No Clinical Signs, Symptoms or Conditions	A0203	Defective Device	13/04/2022	13/04/2022
									A0705	Battery Problem	13/04/2022	13/04/2022
									A070504	Premature Discharge of Battery	13/04/2022	13/04/2022
							F08	Hospitalization or Prolonged Hospitalization	A0203	Defective Device	13/04/2022	13/04/2022
									A0705	Battery Problem	13/04/2022	13/04/2022
									A070504	Premature Discharge of Battery	13/04/2022	13/04/2022
							F1905	Device Revision or Replacement	A0203	Defective Device	13/04/2022	13/04/2022
									A0705	Battery Problem	13/04/2022	13/04/2022
									A070504	Premature Discharge of Battery	13/04/2022	13/04/2022
					CT983	Surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A0203	Defective Device	13/04/2022	13/04/2022
									A0705	Battery Problem	13/04/2022	13/04/2022
									A070504	Premature Discharge of Battery	13/04/2022	13/04/2022
							F08	Hospitalization or Prolonged Hospitalization	A0203	Defective Device	13/04/2022	13/04/2022
									A0705	Battery Problem	13/04/2022	13/04/2022
									A070504	Premature Discharge of Battery	13/04/2022	13/04/2022
							F1905	Device Revision or Replacement	A0203	Defective Device	13/04/2022	13/04/2022
									A0705	Battery Problem	13/04/2022	13/04/2022
									A070504	Premature Discharge of Battery	13/04/2022	13/04/2022
6383	2022	All other reportable incidents	36241	Pacemaker/defibrillator lead	CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A072201	High Impedance		
							F26	No Health Consequences or Impact	A072201	High Impedance		
6384	2022	All other reportable incidents	36241	Pacemaker/defibrillator lead	CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A072201	High Impedance		
							F26	No Health Consequences or Impact	A072201	High Impedance		
6385	2022	All other reportable incidents	37265	Dual-chamber implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	12/04/2022	12/04/2022
							F1901	Additional Surgery	A070504	Premature Discharge of Battery	12/04/2022	12/04/2022
6386	2022	All other reportable incidents	35223	Endocardial/interventricular septal pacing lead	CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A051201	Device Dislodged or Dislocated	01/04/2022	01/04/2022
							F1901	Additional Surgery	A051201	Device Dislodged or Dislocated	01/04/2022	01/04/2022
6387	2021	All other reportable incidents	47270	Cardiac resynchronization therapy implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E0601	Arrhythmia	A070909	Oversensing		
									A071204	No Pacing		
									A12	Connection Problem		
							E2403	No Clinical Signs, Symptoms or Conditions	A070909	Oversensing		
									A071204	No Pacing		
									A12	Connection Problem		
							F1901	Additional Surgery	A070909	Oversensing		
									A071204	No Pacing		
									A12	Connection Problem		
6388	2021	All other reportable incidents	35223	Endocardial/interventricular septal pacing lead	CT1684	Cardiac pacemakers and associated devices	E0601	Arrhythmia	A070909	Oversensing		
									A12	Connection Problem		
							E2403	No Clinical Signs, Symptoms or Conditions	A070909	Oversensing		
									A12	Connection Problem		
							F1901	Additional Surgery	A070909	Oversensing		
									A12	Connection Problem		
6389	2022	All other reportable incidents	47265	Dual-chamber implantable pacemaker, rate-responsive	CT141	Electrodes/Leads and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A1605	Reset Problem		
							F22	Unexpected Diagnostic Intervention	A1605	Reset Problem		
							F26	No Health Consequences or Impact	A1605	Reset Problem		
					CT145	Cardiovascular devices	E2403	No Clinical Signs, Symptoms or Conditions	A1605	Reset Problem		
							F22	Unexpected Diagnostic Intervention	A1605	Reset Problem		
							F26	No Health Consequences or Impact	A1605	Reset Problem		
					CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A1605	Reset Problem		
							F22	Unexpected Diagnostic Intervention	A1605	Reset Problem		
							F26	No Health Consequences or Impact	A1605	Reset Problem		
					CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A1605	Reset Problem		
							F22	Unexpected Diagnostic Intervention	A1605	Reset Problem		
							F26	No Health Consequences or Impact	A1605	Reset Problem		
					CT233	Surgical	E2403	No Clinical Signs, Symptoms or Conditions	A1605	Reset Problem		
							F22	Unexpected Diagnostic Intervention	A1605	Reset Problem		
							F26	No Health Consequences or Impact	A1605	Reset Problem		
					CT2406	Implantable	E2403	No Clinical Signs, Symptoms or Conditions	A1605	Reset Problem		
							F22	Unexpected Diagnostic Intervention	A1605	Reset Problem		
							F26	No Health Consequences or Impact	A1605	Reset Problem		
					CT2407	Active implantable	E2403	No Clinical Signs, Symptoms or Conditions	A1605	Reset Problem		
							F22	Unexpected Diagnostic Intervention	A1605	Reset Problem		
							F26	No Health Consequences or Impact	A1605	Reset Problem		
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A1605	Reset Problem		
							F22	Unexpected Diagnostic Intervention	A1605	Reset Problem		
							F26	No Health Consequences or Impact	A1605	Reset Problem		
					CT2727	Active implant devices/ancillaries	E2403	No Clinical Signs, Symptoms or Conditions	A1605	Reset Problem		
							F22	Unexpected Diagnostic Intervention	A1605	Reset Problem		
							F26	No Health Consequences or Impact	A1605	Reset Problem		
					CT2750	Implantable cardiac electrical stimulators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A1605	Reset Problem		
							F22	Unexpected Diagnostic Intervention	A1605	Reset Problem		
							F26	No Health Consequences or Impact	A1605	Reset Problem		
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A1605	Reset Problem		

							F22	Unexpected Diagnostic Intervention	A1605	Reset Problem		
							F26	No Health Consequences or Impact	A1605	Reset Problem		
					CT2985	Active	E2403	No Clinical Signs, Symptoms or Conditions	A1605	Reset Problem		
							F22	Unexpected Diagnostic Intervention	A1605	Reset Problem		
							F26	No Health Consequences or Impact	A1605	Reset Problem		
					CT3001	Implantable metal	E2403	No Clinical Signs, Symptoms or Conditions	A1605	Reset Problem		
							F22	Unexpected Diagnostic Intervention	A1605	Reset Problem		
							F26	No Health Consequences or Impact	A1605	Reset Problem		
					CT3002	Implantable synthetic polymer	E2403	No Clinical Signs, Symptoms or Conditions	A1605	Reset Problem		
							F22	Unexpected Diagnostic Intervention	A1605	Reset Problem		
							F26	No Health Consequences or Impact	A1605	Reset Problem		
					CT3006	Implantable inorganic material	E2403	No Clinical Signs, Symptoms or Conditions	A1605	Reset Problem		
							F22	Unexpected Diagnostic Intervention	A1605	Reset Problem		
							F26	No Health Consequences or Impact	A1605	Reset Problem		
					CT314	Electrically powered	E2403	No Clinical Signs, Symptoms or Conditions	A1605	Reset Problem		
							F22	Unexpected Diagnostic Intervention	A1605	Reset Problem		
							F26	No Health Consequences or Impact	A1605	Reset Problem		
					CT321	Long-term surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A1605	Reset Problem		
							F22	Unexpected Diagnostic Intervention	A1605	Reset Problem		
							F26	No Health Consequences or Impact	A1605	Reset Problem		
					CT449	Generators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A1605	Reset Problem		
							F22	Unexpected Diagnostic Intervention	A1605	Reset Problem		
							F26	No Health Consequences or Impact	A1605	Reset Problem		
					CT981	Single use	E2403	No Clinical Signs, Symptoms or Conditions	A1605	Reset Problem		
							F22	Unexpected Diagnostic Intervention	A1605	Reset Problem		
							F26	No Health Consequences or Impact	A1605	Reset Problem		
					CT983	Surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A1605	Reset Problem		
							F22	Unexpected Diagnostic Intervention	A1605	Reset Problem		
							F26	No Health Consequences or Impact	A1605	Reset Problem		
6390	2022	All other reportable incidents	17882	Physiologic-monitoring defibrillation system	CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070803	Failure to Power Up		
							F27	Problem Identified During Non-Clinical Procedure	A070803	Failure to Power Up		
6391	2022	All other reportable incidents	35223	Endocardial/interventricular septal pacing lead	CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A07	Electrical /Electronic Property Problem		
									A0701	Capturing Problem		
									A0722	Impedance Problem		
							F26	No Health Consequences or Impact	A07	Electrical /Electronic Property Problem		
									A0701	Capturing Problem		
									A0722	Impedance Problem		
6392	2022	All other reportable incidents	35852	Single-chamber implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem	13/07/2022	13/07/2022
									A070504	Premature Discharge of Battery	13/07/2022	13/07/2022
							F08	Hospitalization or Prolonged Hospitalization	A0705	Battery Problem	13/07/2022	13/07/2022
									A070504	Premature Discharge of Battery	13/07/2022	13/07/2022
							F1903	Device Explantation	A0705	Battery Problem	13/07/2022	13/07/2022
									A070504	Premature Discharge of Battery	13/07/2022	13/07/2022
6393	2022	All other reportable incidents	47265	Dual-chamber implantable pacemaker, rate-responsive	CT141	Electrodes/Leads and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem	11/05/2022	11/05/2022
									A070504	Premature Discharge of Battery	11/05/2022	11/05/2022
							F08	Hospitalization or Prolonged Hospitalization	A0705	Battery Problem	11/05/2022	11/05/2022
									A070504	Premature Discharge of Battery	11/05/2022	11/05/2022
							F1905	Device Revision or Replacement	A0705	Battery Problem	11/05/2022	11/05/2022
									A070504	Premature Discharge of Battery	11/05/2022	11/05/2022
					CT145	Cardiovascular devices	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem	11/05/2022	11/05/2022
									A070504	Premature Discharge of Battery	11/05/2022	11/05/2022
							F08	Hospitalization or Prolonged Hospitalization	A0705	Battery Problem	11/05/2022	11/05/2022
									A070504	Premature Discharge of Battery	11/05/2022	11/05/2022
							F1905	Device Revision or Replacement	A0705	Battery Problem	11/05/2022	11/05/2022
									A070504	Premature Discharge of Battery	11/05/2022	11/05/2022
					CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem	11/05/2022	11/05/2022
									A070504	Premature Discharge of Battery	11/05/2022	11/05/2022
							F08	Hospitalization or Prolonged Hospitalization	A0705	Battery Problem	11/05/2022	11/05/2022
									A070504	Premature Discharge of Battery	11/05/2022	11/05/2022
							F1905	Device Revision or Replacement	A0705	Battery Problem	11/05/2022	11/05/2022
									A070504	Premature Discharge of Battery	11/05/2022	11/05/2022
					CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem	11/05/2022	11/05/2022
									A070504	Premature Discharge of Battery	11/05/2022	11/05/2022
							F08	Hospitalization or Prolonged Hospitalization	A0705	Battery Problem	11/05/2022	11/05/2022
									A070504	Premature Discharge of Battery	11/05/2022	11/05/2022
							F1905	Device Revision or Replacement	A0705	Battery Problem	11/05/2022	11/05/2022
									A070504	Premature Discharge of Battery	11/05/2022	11/05/2022
					CT233	Surgical	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem	11/05/2022	11/05/2022
									A070504	Premature Discharge of Battery	11/05/2022	11/05/2022
							F08	Hospitalization or Prolonged Hospitalization	A0705	Battery Problem	11/05/2022	11/05/2022
									A070504	Premature Discharge of Battery	11/05/2022	11/05/2022
							F1905	Device Revision or Replacement	A0705	Battery Problem	11/05/2022	11/05/2022
									A070504	Premature Discharge of Battery	11/05/2022	11/05/2022
					CT2406	Implantable	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem	11/05/2022	11/05/2022
									A070504	Premature Discharge of Battery	11/05/2022	11/05/2022
							F08	Hospitalization or Prolonged Hospitalization	A0705	Battery Problem	11/05/2022	11/05/2022
									A070504	Premature Discharge of Battery	11/05/2022	11/05/2022
							F1905	Device Revision or Replacement	A0705	Battery Problem	11/05/2022	11/05/2022
									A070504	Premature Discharge of Battery	11/05/2022	11/05/2022
					CT2407	Active implantable	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem	11/05/2022	11/05/2022
									A070504	Premature Discharge of Battery	11/05/2022	11/05/2022
							F08	Hospitalization or Prolonged Hospitalization	A0705	Battery Problem	11/05/2022	11/05/2022
									A070504	Premature Discharge of Battery	11/05/2022	11/05/2022
							F1905	Device Revision or Replacement	A0705	Battery Problem	11/05/2022	11/05/2022
									A070504	Premature Discharge of Battery	11/05/2022	11/05/2022
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem	11/05/2022	11/05/2022
									A070504	Premature Discharge of Battery	11/05/2022	11/05/2022
							F08	Hospitalization or Prolonged Hospitalization	A0705	Battery Problem	11/05/2022	11/05/2022

									A070504	Premature Discharge of Battery	11/05/2022	11/05/2022
							F1905	Device Revision or Replacement	A0705	Battery Problem	11/05/2022	11/05/2022
									A070504	Premature Discharge of Battery	11/05/2022	11/05/2022
					CT2727	Active implant devices/ancillaries	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem	11/05/2022	11/05/2022
									A070504	Premature Discharge of Battery	11/05/2022	11/05/2022
							F08	Hospitalization or Prolonged Hospitalization	A0705	Battery Problem	11/05/2022	11/05/2022
									A070504	Premature Discharge of Battery	11/05/2022	11/05/2022
							F1905	Device Revision or Replacement	A0705	Battery Problem	11/05/2022	11/05/2022
									A070504	Premature Discharge of Battery	11/05/2022	11/05/2022
					CT2750	Implantable cardiac electrical stimulators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem	11/05/2022	11/05/2022
									A070504	Premature Discharge of Battery	11/05/2022	11/05/2022
							F08	Hospitalization or Prolonged Hospitalization	A0705	Battery Problem	11/05/2022	11/05/2022
									A070504	Premature Discharge of Battery	11/05/2022	11/05/2022
							F1905	Device Revision or Replacement	A0705	Battery Problem	11/05/2022	11/05/2022
									A070504	Premature Discharge of Battery	11/05/2022	11/05/2022
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem	11/05/2022	11/05/2022
									A070504	Premature Discharge of Battery	11/05/2022	11/05/2022
							F08	Hospitalization or Prolonged Hospitalization	A0705	Battery Problem	11/05/2022	11/05/2022
									A070504	Premature Discharge of Battery	11/05/2022	11/05/2022
							F1905	Device Revision or Replacement	A0705	Battery Problem	11/05/2022	11/05/2022
									A070504	Premature Discharge of Battery	11/05/2022	11/05/2022
					CT2985	Active	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem	11/05/2022	11/05/2022
									A070504	Premature Discharge of Battery	11/05/2022	11/05/2022
							F08	Hospitalization or Prolonged Hospitalization	A0705	Battery Problem	11/05/2022	11/05/2022
									A070504	Premature Discharge of Battery	11/05/2022	11/05/2022
							F1905	Device Revision or Replacement	A0705	Battery Problem	11/05/2022	11/05/2022
									A070504	Premature Discharge of Battery	11/05/2022	11/05/2022
					CT3001	Implantable metal	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem	11/05/2022	11/05/2022
									A070504	Premature Discharge of Battery	11/05/2022	11/05/2022
							F08	Hospitalization or Prolonged Hospitalization	A0705	Battery Problem	11/05/2022	11/05/2022
									A070504	Premature Discharge of Battery	11/05/2022	11/05/2022
							F1905	Device Revision or Replacement	A0705	Battery Problem	11/05/2022	11/05/2022
									A070504	Premature Discharge of Battery	11/05/2022	11/05/2022
					CT3002	Implantable synthetic polymer	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem	11/05/2022	11/05/2022
									A070504	Premature Discharge of Battery	11/05/2022	11/05/2022
							F08	Hospitalization or Prolonged Hospitalization	A0705	Battery Problem	11/05/2022	11/05/2022
									A070504	Premature Discharge of Battery	11/05/2022	11/05/2022
							F1905	Device Revision or Replacement	A0705	Battery Problem	11/05/2022	11/05/2022
									A070504	Premature Discharge of Battery	11/05/2022	11/05/2022
					CT3006	Implantable inorganic material	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem	11/05/2022	11/05/2022
									A070504	Premature Discharge of Battery	11/05/2022	11/05/2022
							F08	Hospitalization or Prolonged Hospitalization	A0705	Battery Problem	11/05/2022	11/05/2022
									A070504	Premature Discharge of Battery	11/05/2022	11/05/2022
							F1905	Device Revision or Replacement	A0705	Battery Problem	11/05/2022	11/05/2022
									A070504	Premature Discharge of Battery	11/05/2022	11/05/2022
					CT314	Electrically powered	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem	11/05/2022	11/05/2022
									A070504	Premature Discharge of Battery	11/05/2022	11/05/2022
							F08	Hospitalization or Prolonged Hospitalization	A0705	Battery Problem	11/05/2022	11/05/2022
									A070504	Premature Discharge of Battery	11/05/2022	11/05/2022
							F1905	Device Revision or Replacement	A0705	Battery Problem	11/05/2022	11/05/2022
									A070504	Premature Discharge of Battery	11/05/2022	11/05/2022
					CT321	Long-term surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem	11/05/2022	11/05/2022
									A070504	Premature Discharge of Battery	11/05/2022	11/05/2022
							F08	Hospitalization or Prolonged Hospitalization	A0705	Battery Problem	11/05/2022	11/05/2022
									A070504	Premature Discharge of Battery	11/05/2022	11/05/2022
							F1905	Device Revision or Replacement	A0705	Battery Problem	11/05/2022	11/05/2022
									A070504	Premature Discharge of Battery	11/05/2022	11/05/2022
					CT449	Generators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem	11/05/2022	11/05/2022
									A070504	Premature Discharge of Battery	11/05/2022	11/05/2022
							F08	Hospitalization or Prolonged Hospitalization	A0705	Battery Problem	11/05/2022	11/05/2022
									A070504	Premature Discharge of Battery	11/05/2022	11/05/2022
							F1905	Device Revision or Replacement	A0705	Battery Problem	11/05/2022	11/05/2022
									A070504	Premature Discharge of Battery	11/05/2022	11/05/2022
					CT981	Single use	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem	11/05/2022	11/05/2022
									A070504	Premature Discharge of Battery	11/05/2022	11/05/2022
							F08	Hospitalization or Prolonged Hospitalization	A0705	Battery Problem	11/05/2022	11/05/2022
									A070504	Premature Discharge of Battery	11/05/2022	11/05/2022
							F1905	Device Revision or Replacement	A0705	Battery Problem	11/05/2022	11/05/2022
									A070504	Premature Discharge of Battery	11/05/2022	11/05/2022
					CT983	Surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem	11/05/2022	11/05/2022
									A070504	Premature Discharge of Battery	11/05/2022	11/05/2022
							F08	Hospitalization or Prolonged Hospitalization	A0705	Battery Problem	11/05/2022	11/05/2022
									A070504	Premature Discharge of Battery	11/05/2022	11/05/2022
							F1905	Device Revision or Replacement	A0705	Battery Problem	11/05/2022	11/05/2022
									A070504	Premature Discharge of Battery	11/05/2022	11/05/2022
6394	2022	All other reportable incidents	63188	Left-ventricular wireless pacing system sensor/transmitter	CT141	Electrodes/Leads and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070801	Complete Loss of Power	24/06/2022	24/06/2022
					CT145	Cardiovascular devices	E2403	No Clinical Signs, Symptoms or Conditions	A070801	Complete Loss of Power	24/06/2022	24/06/2022
					CT1577	Wireless monitoring system transmitters	E2403	No Clinical Signs, Symptoms or Conditions	A070801	Complete Loss of Power	24/06/2022	24/06/2022
					CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070801	Complete Loss of Power	24/06/2022	24/06/2022
					CT2317	Transmitters/Receivers/Transceivers	E2403	No Clinical Signs, Symptoms or Conditions	A070801	Complete Loss of Power	24/06/2022	24/06/2022
					CT233	Surgical	E2403	No Clinical Signs, Symptoms or Conditions	A070801	Complete Loss of Power	24/06/2022	24/06/2022
					CT2406	Implantable	E2403	No Clinical Signs, Symptoms or Conditions	A070801	Complete Loss of Power	24/06/2022	24/06/2022
					CT2407	Active implantable	E2403	No Clinical Signs, Symptoms or Conditions	A070801	Complete Loss of Power	24/06/2022	24/06/2022
					CT255	Transducers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070801	Complete Loss of Power	24/06/2022	24/06/2022
					CT2726	Physiological monitoring/measuring devices	E2403	No Clinical Signs, Symptoms or Conditions	A070801	Complete Loss of Power	24/06/2022	24/06/2022
					CT2727	Active implant devices/ancillaries	E2403	No Clinical Signs, Symptoms or Conditions	A070801	Complete Loss of Power	24/06/2022	24/06/2022
					CT2750	Implantable cardiac electrical stimulators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070801	Complete Loss of Power	24/06/2022	24/06/2022

					CT2885	Physiological monitoring/measuring system transmitters/receivers/transceivers	E2403	No Clinical Signs, Symptoms or Conditions	A070801	Complete Loss of Power	24/06/2022	24/06/2022
					CT2985	Active	E2403	No Clinical Signs, Symptoms or Conditions	A070801	Complete Loss of Power	24/06/2022	24/06/2022
					CT3001	Implantable metal	E2403	No Clinical Signs, Symptoms or Conditions	A070801	Complete Loss of Power	24/06/2022	24/06/2022
					CT3002	Implantable synthetic polymer	E2403	No Clinical Signs, Symptoms or Conditions	A070801	Complete Loss of Power	24/06/2022	24/06/2022
					CT3006	Implantable inorganic material	E2403	No Clinical Signs, Symptoms or Conditions	A070801	Complete Loss of Power	24/06/2022	24/06/2022
					CT314	Electrically powered	E2403	No Clinical Signs, Symptoms or Conditions	A070801	Complete Loss of Power	24/06/2022	24/06/2022
					CT321	Long-term surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A070801	Complete Loss of Power	24/06/2022	24/06/2022
					CT384	Sensors and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070801	Complete Loss of Power	24/06/2022	24/06/2022
					CT449	Generators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070801	Complete Loss of Power	24/06/2022	24/06/2022
					CT981	Single use	E2403	No Clinical Signs, Symptoms or Conditions	A070801	Complete Loss of Power	24/06/2022	24/06/2022
					CT983	Surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A070801	Complete Loss of Power	24/06/2022	24/06/2022
					CT994	Sensors	E2403	No Clinical Signs, Symptoms or Conditions	A070801	Complete Loss of Power	24/06/2022	24/06/2022
6395	2021	All other reportable incidents	47265	Dual-chamber implantable pacemaker, rate-responsive	CT141	Electrodes/Leads and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A160601	Premature Elective Replacement Indicator	30/07/2021	30/07/2021
							F1903	Device Explantation	A160601	Premature Elective Replacement Indicator	30/07/2021	30/07/2021
					CT145	Cardiovascular devices	E2403	No Clinical Signs, Symptoms or Conditions	A160601	Premature Elective Replacement Indicator	30/07/2021	30/07/2021
							F1903	Device Explantation	A160601	Premature Elective Replacement Indicator	30/07/2021	30/07/2021
					CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A160601	Premature Elective Replacement Indicator	30/07/2021	30/07/2021
							F1903	Device Explantation	A160601	Premature Elective Replacement Indicator	30/07/2021	30/07/2021
					CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A160601	Premature Elective Replacement Indicator	30/07/2021	30/07/2021
							F1903	Device Explantation	A160601	Premature Elective Replacement Indicator	30/07/2021	30/07/2021
					CT233	Surgical	E2403	No Clinical Signs, Symptoms or Conditions	A160601	Premature Elective Replacement Indicator	30/07/2021	30/07/2021
							F1903	Device Explantation	A160601	Premature Elective Replacement Indicator	30/07/2021	30/07/2021
					CT2406	Implantable	E2403	No Clinical Signs, Symptoms or Conditions	A160601	Premature Elective Replacement Indicator	30/07/2021	30/07/2021
							F1903	Device Explantation	A160601	Premature Elective Replacement Indicator	30/07/2021	30/07/2021
					CT2407	Active implantable	E2403	No Clinical Signs, Symptoms or Conditions	A160601	Premature Elective Replacement Indicator	30/07/2021	30/07/2021
							F1903	Device Explantation	A160601	Premature Elective Replacement Indicator	30/07/2021	30/07/2021
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A160601	Premature Elective Replacement Indicator	30/07/2021	30/07/2021
							F1903	Device Explantation	A160601	Premature Elective Replacement Indicator	30/07/2021	30/07/2021
					CT2727	Active implant devices/ancillaries	E2403	No Clinical Signs, Symptoms or Conditions	A160601	Premature Elective Replacement Indicator	30/07/2021	30/07/2021
							F1903	Device Explantation	A160601	Premature Elective Replacement Indicator	30/07/2021	30/07/2021
					CT2750	Implantable cardiac electrical stimulators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A160601	Premature Elective Replacement Indicator	30/07/2021	30/07/2021
							F1903	Device Explantation	A160601	Premature Elective Replacement Indicator	30/07/2021	30/07/2021
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A160601	Premature Elective Replacement Indicator	30/07/2021	30/07/2021
							F1903	Device Explantation	A160601	Premature Elective Replacement Indicator	30/07/2021	30/07/2021
					CT2985	Active	E2403	No Clinical Signs, Symptoms or Conditions	A160601	Premature Elective Replacement Indicator	30/07/2021	30/07/2021
							F1903	Device Explantation	A160601	Premature Elective Replacement Indicator	30/07/2021	30/07/2021
					CT3001	Implantable metal	E2403	No Clinical Signs, Symptoms or Conditions	A160601	Premature Elective Replacement Indicator	30/07/2021	30/07/2021
							F1903	Device Explantation	A160601	Premature Elective Replacement Indicator	30/07/2021	30/07/2021
					CT3002	Implantable synthetic polymer	E2403	No Clinical Signs, Symptoms or Conditions	A160601	Premature Elective Replacement Indicator	30/07/2021	30/07/2021
							F1903	Device Explantation	A160601	Premature Elective Replacement Indicator	30/07/2021	30/07/2021
					CT3006	Implantable inorganic material	E2403	No Clinical Signs, Symptoms or Conditions	A160601	Premature Elective Replacement Indicator	30/07/2021	30/07/2021
							F1903	Device Explantation	A160601	Premature Elective Replacement Indicator	30/07/2021	30/07/2021
					CT314	Electrically powered	E2403	No Clinical Signs, Symptoms or Conditions	A160601	Premature Elective Replacement Indicator	30/07/2021	30/07/2021
							F1903	Device Explantation	A160601	Premature Elective Replacement Indicator	30/07/2021	30/07/2021
					CT321	Long-term surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A160601	Premature Elective Replacement Indicator	30/07/2021	30/07/2021
							F1903	Device Explantation	A160601	Premature Elective Replacement Indicator	30/07/2021	30/07/2021
					CT449	Generators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A160601	Premature Elective Replacement Indicator	30/07/2021	30/07/2021
							F1903	Device Explantation	A160601	Premature Elective Replacement Indicator	30/07/2021	30/07/2021
					CT981	Single use	E2403	No Clinical Signs, Symptoms or Conditions	A160601	Premature Elective Replacement Indicator	30/07/2021	30/07/2021
							F1903	Device Explantation	A160601	Premature Elective Replacement Indicator	30/07/2021	30/07/2021
					CT983	Surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A160601	Premature Elective Replacement Indicator	30/07/2021	30/07/2021
							F1903	Device Explantation	A160601	Premature Elective Replacement Indicator	30/07/2021	30/07/2021
6396	2022	All other reportable incidents	47265	Dual-chamber implantable pacemaker, rate-responsive	CT141	Electrodes/Leads and associated devices	E2312	Fatigue	A070101	Failure to Capture		
									A070102	High Capture Threshold		
									A0712	Pacing Problem		
									A072201	High Impedance		
									A090801	Signal Artifact/Noise		
							E2403	No Clinical Signs, Symptoms or Conditions	A070101	Failure to Capture		
									A070102	High Capture Threshold		
									A0712	Pacing Problem		
									A072201	High Impedance		
									A090801	Signal Artifact/Noise		
							F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A070101	Failure to Capture		
									A070102	High Capture Threshold		
									A0712	Pacing Problem		
									A072201	High Impedance		
									A090801	Signal Artifact/Noise		
					CT145	Cardiovascular devices	E2312	Fatigue	A070101	Failure to Capture		
									A070102	High Capture Threshold		
									A0712	Pacing Problem		
									A072201	High Impedance		
									A090801	Signal Artifact/Noise		
							E2403	No Clinical Signs, Symptoms or Conditions	A070101	Failure to Capture		
									A070102	High Capture Threshold		
									A0712	Pacing Problem		
									A072201	High Impedance		
									A090801	Signal Artifact/Noise		
							F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A070101	Failure to Capture		
									A070102	High Capture Threshold		
									A0712	Pacing Problem		
									A072201	High Impedance		
									A090801	Signal Artifact/Noise		
					CT1684	Cardiac pacemakers and associated devices	E2312	Fatigue	A070101	Failure to Capture		
									A070102	High Capture Threshold		
									A0712	Pacing Problem		
									A072201	High Impedance		
									A090801	Signal Artifact/Noise		

							E2403	No Clinical Signs, Symptoms or Conditions	A070101	Failure to Capture		
									A070102	High Capture Threshold		
									A0712	Pacing Problem		
									A072201	High Impedance		
									A090801	Signal Artifact/Noise		
							F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A070101	Failure to Capture		
									A070102	High Capture Threshold		
									A0712	Pacing Problem		
									A072201	High Impedance		
									A090801	Signal Artifact/Noise		
					CT1685	Cardiac pacemakers	E2312	Fatigue	A070101	Failure to Capture		
									A070102	High Capture Threshold		
									A0712	Pacing Problem		
									A072201	High Impedance		
									A090801	Signal Artifact/Noise		
							E2403	No Clinical Signs, Symptoms or Conditions	A070101	Failure to Capture		
									A070102	High Capture Threshold		
									A0712	Pacing Problem		
									A072201	High Impedance		
									A090801	Signal Artifact/Noise		
							F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A070101	Failure to Capture		
									A070102	High Capture Threshold		
									A0712	Pacing Problem		
									A072201	High Impedance		
									A090801	Signal Artifact/Noise		
					CT233	Surgical	E2312	Fatigue	A070101	Failure to Capture		
									A070102	High Capture Threshold		
									A0712	Pacing Problem		
									A072201	High Impedance		
									A090801	Signal Artifact/Noise		
							E2403	No Clinical Signs, Symptoms or Conditions	A070101	Failure to Capture		
									A070102	High Capture Threshold		
									A0712	Pacing Problem		
									A072201	High Impedance		
									A090801	Signal Artifact/Noise		
							F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A070101	Failure to Capture		
									A070102	High Capture Threshold		
									A0712	Pacing Problem		
									A072201	High Impedance		
									A090801	Signal Artifact/Noise		
					CT2406	Implantable	E2312	Fatigue	A070101	Failure to Capture		
									A070102	High Capture Threshold		
									A0712	Pacing Problem		
									A072201	High Impedance		
									A090801	Signal Artifact/Noise		
							E2403	No Clinical Signs, Symptoms or Conditions	A070101	Failure to Capture		
									A070102	High Capture Threshold		
									A0712	Pacing Problem		
									A072201	High Impedance		
									A090801	Signal Artifact/Noise		
							F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A070101	Failure to Capture		
									A070102	High Capture Threshold		
									A0712	Pacing Problem		
									A072201	High Impedance		
									A090801	Signal Artifact/Noise		
					CT2407	Active implantable	E2312	Fatigue	A070101	Failure to Capture		
									A070102	High Capture Threshold		
									A0712	Pacing Problem		
									A072201	High Impedance		
									A090801	Signal Artifact/Noise		
							E2403	No Clinical Signs, Symptoms or Conditions	A070101	Failure to Capture		
									A070102	High Capture Threshold		
									A0712	Pacing Problem		
									A072201	High Impedance		
									A090801	Signal Artifact/Noise		
							F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A070101	Failure to Capture		
									A070102	High Capture Threshold		
									A0712	Pacing Problem		
									A072201	High Impedance		
									A090801	Signal Artifact/Noise		
					CT257	Implantable cardiac pacemakers	E2312	Fatigue	A070101	Failure to Capture		
									A070102	High Capture Threshold		
									A0712	Pacing Problem		
									A072201	High Impedance		
									A090801	Signal Artifact/Noise		
							E2403	No Clinical Signs, Symptoms or Conditions	A070101	Failure to Capture		
									A070102	High Capture Threshold		
									A0712	Pacing Problem		
									A072201	High Impedance		
									A090801	Signal Artifact/Noise		
							F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A070101	Failure to Capture		
									A070102	High Capture Threshold		
									A0712	Pacing Problem		
									A072201	High Impedance		
									A090801	Signal Artifact/Noise		
					CT2727	Active implant devices/ancillaries	E2312	Fatigue	A070101	Failure to Capture		
									A070102	High Capture Threshold		
									A0712	Pacing Problem		

									A072201	High Impedance		
									A090801	Signal Artifact/Noise		
							E2403	No Clinical Signs, Symptoms or Conditions	A070101	Failure to Capture		
									A070102	High Capture Threshold		
									A0712	Pacing Problem		
									A072201	High Impedance		
									A090801	Signal Artifact/Noise		
							F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A070101	Failure to Capture		
									A070102	High Capture Threshold		
									A0712	Pacing Problem		
									A072201	High Impedance		
									A090801	Signal Artifact/Noise		
					CT2750	Implantable cardiac electrical stimulators and associated devices	E2312	Fatigue	A070101	Failure to Capture		
									A070102	High Capture Threshold		
									A0712	Pacing Problem		
									A072201	High Impedance		
									A090801	Signal Artifact/Noise		
							E2403	No Clinical Signs, Symptoms or Conditions	A070101	Failure to Capture		
									A070102	High Capture Threshold		
									A0712	Pacing Problem		
									A072201	High Impedance		
									A090801	Signal Artifact/Noise		
							F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A070101	Failure to Capture		
									A070102	High Capture Threshold		
									A0712	Pacing Problem		
									A072201	High Impedance		
									A090801	Signal Artifact/Noise		
					CT2866	Implantable cardiac electrical stimulators	E2312	Fatigue	A070101	Failure to Capture		
									A070102	High Capture Threshold		
									A0712	Pacing Problem		
									A072201	High Impedance		
									A090801	Signal Artifact/Noise		
							E2403	No Clinical Signs, Symptoms or Conditions	A070101	Failure to Capture		
									A070102	High Capture Threshold		
									A0712	Pacing Problem		
									A072201	High Impedance		
									A090801	Signal Artifact/Noise		
							F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A070101	Failure to Capture		
									A070102	High Capture Threshold		
									A0712	Pacing Problem		
									A072201	High Impedance		
									A090801	Signal Artifact/Noise		
					CT2985	Active	E2312	Fatigue	A070101	Failure to Capture		
									A070102	High Capture Threshold		
									A0712	Pacing Problem		
									A072201	High Impedance		
									A090801	Signal Artifact/Noise		
							E2403	No Clinical Signs, Symptoms or Conditions	A070101	Failure to Capture		
									A070102	High Capture Threshold		
									A0712	Pacing Problem		
									A072201	High Impedance		
									A090801	Signal Artifact/Noise		
							F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A070101	Failure to Capture		
									A070102	High Capture Threshold		
									A0712	Pacing Problem		
									A072201	High Impedance		
									A090801	Signal Artifact/Noise		
					CT3001	Implantable metal	E2312	Fatigue	A070101	Failure to Capture		
									A070102	High Capture Threshold		
									A0712	Pacing Problem		
									A072201	High Impedance		
									A090801	Signal Artifact/Noise		
							E2403	No Clinical Signs, Symptoms or Conditions	A070101	Failure to Capture		
									A070102	High Capture Threshold		
									A0712	Pacing Problem		
									A072201	High Impedance		
									A090801	Signal Artifact/Noise		
							F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A070101	Failure to Capture		
									A070102	High Capture Threshold		
									A0712	Pacing Problem		
									A072201	High Impedance		
									A090801	Signal Artifact/Noise		
					CT3002	Implantable synthetic polymer	E2312	Fatigue	A070101	Failure to Capture		
									A070102	High Capture Threshold		
									A0712	Pacing Problem		
									A072201	High Impedance		
									A090801	Signal Artifact/Noise		
							E2403	No Clinical Signs, Symptoms or Conditions	A070101	Failure to Capture		
									A070102	High Capture Threshold		
									A0712	Pacing Problem		
									A072201	High Impedance		
									A090801	Signal Artifact/Noise		
							F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A070101	Failure to Capture		
									A070102	High Capture Threshold		
									A0712	Pacing Problem		
									A072201	High Impedance		
									A090801	Signal Artifact/Noise		
					CT3006	Implantable inorganic material	E2312	Fatigue	A070101	Failure to Capture		

									A070102	High Capture Threshold		
									A0712	Pacing Problem		
									A072201	High Impedance		
									A090801	Signal Artifact/Noise		
							E2403	No Clinical Signs, Symptoms or Conditions	A070101	Failure to Capture		
									A070102	High Capture Threshold		
									A0712	Pacing Problem		
									A072201	High Impedance		
									A090801	Signal Artifact/Noise		
							F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A070101	Failure to Capture		
									A070102	High Capture Threshold		
									A0712	Pacing Problem		
									A072201	High Impedance		
									A090801	Signal Artifact/Noise		
					CT314	Electrically powered	E2312	Fatigue	A070101	Failure to Capture		
									A070102	High Capture Threshold		
									A0712	Pacing Problem		
									A072201	High Impedance		
									A090801	Signal Artifact/Noise		
							E2403	No Clinical Signs, Symptoms or Conditions	A070101	Failure to Capture		
									A070102	High Capture Threshold		
									A0712	Pacing Problem		
									A072201	High Impedance		
									A090801	Signal Artifact/Noise		
							F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A070101	Failure to Capture		
									A070102	High Capture Threshold		
									A0712	Pacing Problem		
									A072201	High Impedance		
									A090801	Signal Artifact/Noise		
					CT321	Long-term surgical invasive	E2312	Fatigue	A070101	Failure to Capture		
									A070102	High Capture Threshold		
									A0712	Pacing Problem		
									A072201	High Impedance		
									A090801	Signal Artifact/Noise		
							E2403	No Clinical Signs, Symptoms or Conditions	A070101	Failure to Capture		
									A070102	High Capture Threshold		
									A0712	Pacing Problem		
									A072201	High Impedance		
									A090801	Signal Artifact/Noise		
							F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A070101	Failure to Capture		
									A070102	High Capture Threshold		
									A0712	Pacing Problem		
									A072201	High Impedance		
									A090801	Signal Artifact/Noise		
					CT449	Generators and associated devices	E2312	Fatigue	A070101	Failure to Capture		
									A070102	High Capture Threshold		
									A0712	Pacing Problem		
									A072201	High Impedance		
									A090801	Signal Artifact/Noise		
							E2403	No Clinical Signs, Symptoms or Conditions	A070101	Failure to Capture		
									A070102	High Capture Threshold		
									A0712	Pacing Problem		
									A072201	High Impedance		
									A090801	Signal Artifact/Noise		
							F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A070101	Failure to Capture		
									A070102	High Capture Threshold		
									A0712	Pacing Problem		
									A072201	High Impedance		
									A090801	Signal Artifact/Noise		
					CT981	Single use	E2312	Fatigue	A070101	Failure to Capture		
									A070102	High Capture Threshold		
									A0712	Pacing Problem		
									A072201	High Impedance		
									A090801	Signal Artifact/Noise		
							E2403	No Clinical Signs, Symptoms or Conditions	A070101	Failure to Capture		
									A070102	High Capture Threshold		
									A0712	Pacing Problem		
									A072201	High Impedance		
									A090801	Signal Artifact/Noise		
							F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A070101	Failure to Capture		
									A070102	High Capture Threshold		
									A0712	Pacing Problem		
									A072201	High Impedance		
									A090801	Signal Artifact/Noise		
					CT983	Surgical invasive	E2312	Fatigue	A070101	Failure to Capture		
									A070102	High Capture Threshold		
									A0712	Pacing Problem		
									A072201	High Impedance		
									A090801	Signal Artifact/Noise		
							E2403	No Clinical Signs, Symptoms or Conditions	A070101	Failure to Capture		
									A070102	High Capture Threshold		
									A0712	Pacing Problem		
									A072201	High Impedance		
									A090801	Signal Artifact/Noise		
							F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A070101	Failure to Capture		
									A070102	High Capture Threshold		
									A0712	Pacing Problem		
									A072201	High Impedance		

									A090801	Signal Artifact/Noise			
6397	2021	All other reportable incidents	47265	Dual-chamber implantable pacemaker, rate-responsive	CT141	Electrodes/Leads and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT145	Cardiovascular devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT233	Surgical	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT2406	Implantable	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT2407	Active implantable	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT2727	Active implant devices/ancillaries	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT2750	Implantable cardiac electrical stimulators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT2985	Active	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT3001	Implantable metal	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT3002	Implantable synthetic polymer	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT3006	Implantable inorganic material	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT314	Electrically powered	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT321	Long-term surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT449	Generators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT981	Single use	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT983	Surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
6398	2022	All other reportable incidents	47263	Cardiac resynchronization therapy implantable pacemaker	CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F1901	Additional Surgery	A070504	Premature Discharge of Battery		
					CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F1901	Additional Surgery	A070504	Premature Discharge of Battery		
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F1901	Additional Surgery	A070504	Premature Discharge of Battery		
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F1901	Additional Surgery	A070504	Premature Discharge of Battery		
6399	2021	All other reportable incidents	35223	Endocardial/interventricular septal pacing lead	CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A072202	Low Impedance		
							F26	No Health Consequences or Impact	A072202	Low Impedance		
6400	2021	All other reportable incidents	47265	Dual-chamber implantable pacemaker, rate-responsive	CT141	Electrodes/Leads and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	30/11/2021	30/11/2021
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	30/11/2021	30/11/2021
							F1903	Device Explantation	A070504	Premature Discharge of Battery	30/11/2021	30/11/2021
					CT145	Cardiovascular devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	30/11/2021	30/11/2021
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	30/11/2021	30/11/2021
							F1903	Device Explantation	A070504	Premature Discharge of Battery	30/11/2021	30/11/2021
					CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	30/11/2021	30/11/2021
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	30/11/2021	30/11/2021
							F1903	Device Explantation	A070504	Premature Discharge of Battery	30/11/2021	30/11/2021
					CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	30/11/2021	30/11/2021
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	30/11/2021	30/11/2021
							F1903	Device Explantation	A070504	Premature Discharge of Battery	30/11/2021	30/11/2021
					CT233	Surgical	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	30/11/2021	30/11/2021
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	30/11/2021	30/11/2021
							F1903	Device Explantation	A070504	Premature Discharge of Battery	30/11/2021	30/11/2021
					CT2406	Implantable	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	30/11/2021	30/11/2021
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	30/11/2021	30/11/2021
							F1903	Device Explantation	A070504	Premature Discharge of Battery	30/11/2021	30/11/2021
					CT2407	Active implantable	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	30/11/2021	30/11/2021
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	30/11/2021	30/11/2021
							F1903	Device Explantation	A070504	Premature Discharge of Battery	30/11/2021	30/11/2021
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	30/11/2021	30/11/2021
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	30/11/2021	30/11/2021
							F1903	Device Explantation	A070504	Premature Discharge of Battery	30/11/2021	30/11/2021
					CT2727	Active implant devices/ancillaries	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	30/11/2021	30/11/2021
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	30/11/2021	30/11/2021
							F1903	Device Explantation	A070504	Premature Discharge of Battery	30/11/2021	30/11/2021
					CT2750	Implantable cardiac electrical stimulators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	30/11/2021	30/11/2021
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	30/11/2021	30/11/2021
							F1903	Device Explantation	A070504	Premature Discharge of Battery	30/11/2021	30/11/2021
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	30/11/2021	30/11/2021
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	30/11/2021	30/11/2021
							F1903	Device Explantation	A070504	Premature Discharge of Battery	30/11/2021	30/11/2021
					CT2985	Active	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	30/11/2021	30/11/2021
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	30/11/2021	30/11/2021
							F1903	Device Explantation	A070504	Premature Discharge of Battery	30/11/2021	30/11/2021
					CT3001	Implantable metal	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	30/11/2021	30/11/2021

						F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	30/11/2021	30/11/2021
						F1903	Device Explantation	A070504	Premature Discharge of Battery	30/11/2021	30/11/2021
					CT3002	Implantable synthetic polymer	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	30/11/2021	30/11/2021
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	30/11/2021	30/11/2021
							F1903	Device Explantation	A070504	Premature Discharge of Battery	30/11/2021	30/11/2021
					CT3006	Implantable inorganic material	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	30/11/2021	30/11/2021
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	30/11/2021	30/11/2021
							F1903	Device Explantation	A070504	Premature Discharge of Battery	30/11/2021	30/11/2021
					CT314	Electrically powered	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	30/11/2021	30/11/2021
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	30/11/2021	30/11/2021
							F1903	Device Explantation	A070504	Premature Discharge of Battery	30/11/2021	30/11/2021
					CT321	Long-term surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	30/11/2021	30/11/2021
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	30/11/2021	30/11/2021
							F1903	Device Explantation	A070504	Premature Discharge of Battery	30/11/2021	30/11/2021
					CT449	Generators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	30/11/2021	30/11/2021
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	30/11/2021	30/11/2021
							F1903	Device Explantation	A070504	Premature Discharge of Battery	30/11/2021	30/11/2021
					CT981	Single use	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	30/11/2021	30/11/2021
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	30/11/2021	30/11/2021
							F1903	Device Explantation	A070504	Premature Discharge of Battery	30/11/2021	30/11/2021
					CT983	Surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	30/11/2021	30/11/2021
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	30/11/2021	30/11/2021
							F1903	Device Explantation	A070504	Premature Discharge of Battery	30/11/2021	30/11/2021
6401	2022	All other reportable incidents	37265	Dual-chamber implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
6402	2022	All other reportable incidents	47265	Dual-chamber implantable pacemaker, rate-responsive	CT141	Electrodes/Leads and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	31/05/2022	31/05/2022
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	31/05/2022	31/05/2022
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	31/05/2022	31/05/2022
					CT145	Cardiovascular devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	31/05/2022	31/05/2022
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	31/05/2022	31/05/2022
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	31/05/2022	31/05/2022
					CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	31/05/2022	31/05/2022
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	31/05/2022	31/05/2022
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	31/05/2022	31/05/2022
					CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	31/05/2022	31/05/2022
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	31/05/2022	31/05/2022
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	31/05/2022	31/05/2022
					CT233	Surgical	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	31/05/2022	31/05/2022
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	31/05/2022	31/05/2022
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	31/05/2022	31/05/2022
					CT2406	Implantable	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	31/05/2022	31/05/2022
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	31/05/2022	31/05/2022
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	31/05/2022	31/05/2022
					CT2407	Active implantable	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	31/05/2022	31/05/2022
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	31/05/2022	31/05/2022
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	31/05/2022	31/05/2022
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	31/05/2022	31/05/2022
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	31/05/2022	31/05/2022
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	31/05/2022	31/05/2022
					CT2727	Active implant devices/ancillaries	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	31/05/2022	31/05/2022
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	31/05/2022	31/05/2022
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	31/05/2022	31/05/2022
					CT2750	Implantable cardiac electrical stimulators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	31/05/2022	31/05/2022
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	31/05/2022	31/05/2022
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	31/05/2022	31/05/2022
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	31/05/2022	31/05/2022
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	31/05/2022	31/05/2022
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	31/05/2022	31/05/2022
					CT2985	Active	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	31/05/2022	31/05/2022
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	31/05/2022	31/05/2022
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	31/05/2022	31/05/2022
					CT3001	Implantable metal	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	31/05/2022	31/05/2022
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	31/05/2022	31/05/2022
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	31/05/2022	31/05/2022
					CT3002	Implantable synthetic polymer	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	31/05/2022	31/05/2022
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	31/05/2022	31/05/2022
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	31/05/2022	31/05/2022
					CT3006	Implantable inorganic material	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	31/05/2022	31/05/2022
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	31/05/2022	31/05/2022
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	31/05/2022	31/05/2022
					CT314	Electrically powered	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	31/05/2022	31/05/2022
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	31/05/2022	31/05/2022
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	31/05/2022	31/05/2022
					CT321	Long-term surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	31/05/2022	31/05/2022
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	31/05/2022	31/05/2022
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	31/05/2022	31/05/2022
					CT449	Generators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	31/05/2022	31/05/2022
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	31/05/2022	31/05/2022
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	31/05/2022	31/05/2022
					CT981	Single use	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	31/05/2022	31/05/2022
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	31/05/2022	31/05/2022
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	31/05/2022	31/05/2022
					CT983	Surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	31/05/2022	31/05/2022
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	31/05/2022	31/05/2022
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	31/05/2022	31/05/2022
6403	2021	All other reportable incidents	35852	Single-chamber implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	28/09/2021	28/09/2021
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	28/09/2021	28/09/2021
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	28/09/2021	28/09/2021

6404	2021	All other reportable incidents	47265	Dual-chamber implantable pacemaker, rate-responsive	CT141	Electrodes/Leads and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	04/11/2021	04/11/2021
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	04/11/2021	04/11/2021
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	04/11/2021	04/11/2021
					CT145	Cardiovascular devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	04/11/2021	04/11/2021
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	04/11/2021	04/11/2021
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	04/11/2021	04/11/2021
					CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	04/11/2021	04/11/2021
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	04/11/2021	04/11/2021
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	04/11/2021	04/11/2021
					CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	04/11/2021	04/11/2021
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	04/11/2021	04/11/2021
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	04/11/2021	04/11/2021
					CT233	Surgical	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	04/11/2021	04/11/2021
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	04/11/2021	04/11/2021
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	04/11/2021	04/11/2021
					CT2406	Implantable	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	04/11/2021	04/11/2021
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	04/11/2021	04/11/2021
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	04/11/2021	04/11/2021
					CT2407	Active implantable	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	04/11/2021	04/11/2021
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	04/11/2021	04/11/2021
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	04/11/2021	04/11/2021
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	04/11/2021	04/11/2021
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	04/11/2021	04/11/2021
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	04/11/2021	04/11/2021
					CT2727	Active implant devices/ancillaries	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	04/11/2021	04/11/2021
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	04/11/2021	04/11/2021
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	04/11/2021	04/11/2021
					CT2750	Implantable cardiac electrical stimulators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	04/11/2021	04/11/2021
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	04/11/2021	04/11/2021
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	04/11/2021	04/11/2021
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	04/11/2021	04/11/2021
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	04/11/2021	04/11/2021
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	04/11/2021	04/11/2021
					CT2985	Active	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	04/11/2021	04/11/2021
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	04/11/2021	04/11/2021
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	04/11/2021	04/11/2021
					CT3001	Implantable metal	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	04/11/2021	04/11/2021
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	04/11/2021	04/11/2021
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	04/11/2021	04/11/2021
					CT3002	Implantable synthetic polymer	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	04/11/2021	04/11/2021
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	04/11/2021	04/11/2021
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	04/11/2021	04/11/2021
					CT3006	Implantable inorganic material	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	04/11/2021	04/11/2021
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	04/11/2021	04/11/2021
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	04/11/2021	04/11/2021
					CT314	Electrically powered	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	04/11/2021	04/11/2021
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	04/11/2021	04/11/2021
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	04/11/2021	04/11/2021
					CT321	Long-term surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	04/11/2021	04/11/2021
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	04/11/2021	04/11/2021
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	04/11/2021	04/11/2021
					CT449	Generators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	04/11/2021	04/11/2021
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	04/11/2021	04/11/2021
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	04/11/2021	04/11/2021
					CT981	Single use	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	04/11/2021	04/11/2021
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	04/11/2021	04/11/2021
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	04/11/2021	04/11/2021
					CT983	Surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	04/11/2021	04/11/2021
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	04/11/2021	04/11/2021
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	04/11/2021	04/11/2021
6405	2021	All other reportable incidents	36241	Pacemaker/defibrillator lead	CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A05	Mechanical Problem	05/07/2021	05/07/2021
							F1908	Prolonged Surgery	A05	Mechanical Problem	05/07/2021	05/07/2021
6406	2021	All other reportable incidents	47265	Dual-chamber implantable pacemaker, rate-responsive	CT141	Electrodes/Leads and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	08/09/2021	08/09/2021
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery	08/09/2021	08/09/2021
					CT145	Cardiovascular devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	08/09/2021	08/09/2021
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery	08/09/2021	08/09/2021
					CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	08/09/2021	08/09/2021
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery	08/09/2021	08/09/2021
					CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	08/09/2021	08/09/2021
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery	08/09/2021	08/09/2021
					CT233	Surgical	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	08/09/2021	08/09/2021
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery	08/09/2021	08/09/2021
					CT2406	Implantable	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	08/09/2021	08/09/2021
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery	08/09/2021	08/09/2021
					CT2407	Active implantable	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	08/09/2021	08/09/2021
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery	08/09/2021	08/09/2021
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	08/09/2021	08/09/2021
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery	08/09/2021	08/09/2021
					CT2727	Active implant devices/ancillaries	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	08/09/2021	08/09/2021
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery	08/09/2021	08/09/2021
					CT2750	Implantable cardiac electrical stimulators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	08/09/2021	08/09/2021
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery	08/09/2021	08/09/2021
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	08/09/2021	08/09/2021
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery	08/09/2021	08/09/2021
					CT2985	Active	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	08/09/2021	08/09/2021
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery	08/09/2021	08/09/2021
					CT3001	Implantable metal	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	08/09/2021	08/09/2021
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery	08/09/2021	08/09/2021

					CT3002	Implantable synthetic polymer	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	08/09/2021	08/09/2021
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery	08/09/2021	08/09/2021
					CT3006	Implantable inorganic material	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	08/09/2021	08/09/2021
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery	08/09/2021	08/09/2021
					CT314	Electrically powered	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	08/09/2021	08/09/2021
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery	08/09/2021	08/09/2021
					CT321	Long-term surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	08/09/2021	08/09/2021
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery	08/09/2021	08/09/2021
					CT449	Generators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	08/09/2021	08/09/2021
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery	08/09/2021	08/09/2021
					CT981	Single use	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	08/09/2021	08/09/2021
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery	08/09/2021	08/09/2021
					CT983	Surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	08/09/2021	08/09/2021
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery	08/09/2021	08/09/2021
6407	2021	All other reportable incidents	47265	Dual-chamber implantable pacemaker, rate-responsive	CT141	Electrodes/Leads and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	25/08/2021	25/08/2021
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery	25/08/2021	25/08/2021
					CT145	Cardiovascular devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	25/08/2021	25/08/2021
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery	25/08/2021	25/08/2021
					CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	25/08/2021	25/08/2021
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery	25/08/2021	25/08/2021
					CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	25/08/2021	25/08/2021
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery	25/08/2021	25/08/2021
					CT233	Surgical	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	25/08/2021	25/08/2021
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery	25/08/2021	25/08/2021
					CT2406	Implantable	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	25/08/2021	25/08/2021
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery	25/08/2021	25/08/2021
					CT2407	Active implantable	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	25/08/2021	25/08/2021
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery	25/08/2021	25/08/2021
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	25/08/2021	25/08/2021
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery	25/08/2021	25/08/2021
					CT2727	Active implant devices/ancillaries	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	25/08/2021	25/08/2021
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery	25/08/2021	25/08/2021
					CT2750	Implantable cardiac electrical stimulators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	25/08/2021	25/08/2021
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery	25/08/2021	25/08/2021
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	25/08/2021	25/08/2021
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery	25/08/2021	25/08/2021
					CT2985	Active	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	25/08/2021	25/08/2021
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery	25/08/2021	25/08/2021
					CT3001	Implantable metal	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	25/08/2021	25/08/2021
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery	25/08/2021	25/08/2021
					CT3002	Implantable synthetic polymer	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	25/08/2021	25/08/2021
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery	25/08/2021	25/08/2021
					CT3006	Implantable inorganic material	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	25/08/2021	25/08/2021
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery	25/08/2021	25/08/2021
					CT314	Electrically powered	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	25/08/2021	25/08/2021
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery	25/08/2021	25/08/2021
					CT321	Long-term surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	25/08/2021	25/08/2021
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery	25/08/2021	25/08/2021
					CT449	Generators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	25/08/2021	25/08/2021
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery	25/08/2021	25/08/2021
					CT981	Single use	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	25/08/2021	25/08/2021
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery	25/08/2021	25/08/2021
					CT983	Surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	25/08/2021	25/08/2021
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery	25/08/2021	25/08/2021
6408	2021	All other reportable incidents	35852	Single-chamber implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem	09/09/2021	09/09/2021
									A070504	Premature Discharge of Battery	09/09/2021	09/09/2021
							F08	Hospitalization or Prolonged Hospitalization	A0705	Battery Problem	09/09/2021	09/09/2021
									A070504	Premature Discharge of Battery	09/09/2021	09/09/2021
							F1905	Device Revision or Replacement	A0705	Battery Problem	09/09/2021	09/09/2021
									A070504	Premature Discharge of Battery	09/09/2021	09/09/2021
6409	2021	All other reportable incidents	35852	Single-chamber implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem	19/08/2021	19/08/2021
									A070504	Premature Discharge of Battery	19/08/2021	19/08/2021
							F08	Hospitalization or Prolonged Hospitalization	A0705	Battery Problem	19/08/2021	19/08/2021
									A070504	Premature Discharge of Battery	19/08/2021	19/08/2021
							F1905	Device Revision or Replacement	A0705	Battery Problem	19/08/2021	19/08/2021
									A070504	Premature Discharge of Battery	19/08/2021	19/08/2021
6410	2021	All other reportable incidents	47265	Dual-chamber implantable pacemaker, rate-responsive	CT141	Electrodes/Leads and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	20/08/2021	20/08/2021
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	20/08/2021	20/08/2021
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	20/08/2021	20/08/2021
					CT145	Cardiovascular devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	20/08/2021	20/08/2021
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	20/08/2021	20/08/2021
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	20/08/2021	20/08/2021
					CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	20/08/2021	20/08/2021
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	20/08/2021	20/08/2021
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	20/08/2021	20/08/2021
					CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	20/08/2021	20/08/2021
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	20/08/2021	20/08/2021
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	20/08/2021	20/08/2021
					CT233	Surgical	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	20/08/2021	20/08/2021
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	20/08/2021	20/08/2021
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	20/08/2021	20/08/2021
					CT2406	Implantable	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	20/08/2021	20/08/2021
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	20/08/2021	20/08/2021
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	20/08/2021	20/08/2021
					CT2407	Active implantable	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	20/08/2021	20/08/2021
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	20/08/2021	20/08/2021
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	20/08/2021	20/08/2021
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	20/08/2021	20/08/2021

							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	20/08/2021	20/08/2021
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	20/08/2021	20/08/2021
					CT2727	Active implant devices/ancillaries	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	20/08/2021	20/08/2021
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	20/08/2021	20/08/2021
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	20/08/2021	20/08/2021
					CT2750	Implantable cardiac electrical stimulators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	20/08/2021	20/08/2021
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	20/08/2021	20/08/2021
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	20/08/2021	20/08/2021
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	20/08/2021	20/08/2021
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	20/08/2021	20/08/2021
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	20/08/2021	20/08/2021
					CT2985	Active	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	20/08/2021	20/08/2021
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	20/08/2021	20/08/2021
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	20/08/2021	20/08/2021
					CT3001	Implantable metal	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	20/08/2021	20/08/2021
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	20/08/2021	20/08/2021
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	20/08/2021	20/08/2021
					CT3002	Implantable synthetic polymer	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	20/08/2021	20/08/2021
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	20/08/2021	20/08/2021
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	20/08/2021	20/08/2021
					CT3006	Implantable inorganic material	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	20/08/2021	20/08/2021
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	20/08/2021	20/08/2021
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	20/08/2021	20/08/2021
					CT314	Electrically powered	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	20/08/2021	20/08/2021
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	20/08/2021	20/08/2021
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	20/08/2021	20/08/2021
					CT321	Long-term surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	20/08/2021	20/08/2021
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	20/08/2021	20/08/2021
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	20/08/2021	20/08/2021
					CT449	Generators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	20/08/2021	20/08/2021
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	20/08/2021	20/08/2021
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	20/08/2021	20/08/2021
					CT981	Single use	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	20/08/2021	20/08/2021
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	20/08/2021	20/08/2021
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	20/08/2021	20/08/2021
					CT983	Surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	20/08/2021	20/08/2021
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	20/08/2021	20/08/2021
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	20/08/2021	20/08/2021
6411	2022	All other reportable incidents	37265	Dual-chamber implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A070909	Oversensing		
									A090801	Signal Artifact/Noise		
									A1301	Failure to Read Input Signal		
							F26	No Health Consequences or Impact	A070909	Oversensing		
									A090801	Signal Artifact/Noise		
									A1301	Failure to Read Input Signal		
6412	2022	All other reportable incidents	17882	Physiologic-monitoring defibrillation system	CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070903	Failure to Analyze Signal		
							F26	No Health Consequences or Impact	A070903	Failure to Analyze Signal		
6413	2021	All other reportable incidents	17882	Physiologic-monitoring defibrillation system	CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A090811	Unable to Obtain Readings		
							F27	Problem Identified During Non-Clinical Procedure	A090811	Unable to Obtain Readings		
6414	2022	All other reportable incidents	37265	Dual-chamber implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
6415	2022	All other reportable incidents	47263	Cardiac resynchronization therapy implantable pacemaker	CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	18/03/2022	18/03/2022
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	18/03/2022	18/03/2022
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	18/03/2022	18/03/2022
					CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	18/03/2022	18/03/2022
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	18/03/2022	18/03/2022
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	18/03/2022	18/03/2022
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	18/03/2022	18/03/2022
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	18/03/2022	18/03/2022
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	18/03/2022	18/03/2022
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	18/03/2022	18/03/2022
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	18/03/2022	18/03/2022
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	18/03/2022	18/03/2022
6416	2022	All other reportable incidents	37265	Dual-chamber implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E060110	Ventricular Fibrillation	A070101	Failure to Capture		
									A0712	Pacing Problem		
									A071202	Inaccurate Synchronization		
							E2403	No Clinical Signs, Symptoms or Conditions	A070101	Failure to Capture		
									A0712	Pacing Problem		
									A071202	Inaccurate Synchronization		
							F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A070101	Failure to Capture		
									A0712	Pacing Problem		
									A071202	Inaccurate Synchronization		
6417	2021	All other reportable incidents	35852	Single-chamber implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	26/08/2021	26/08/2021
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	26/08/2021	26/08/2021
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	26/08/2021	26/08/2021
6418	2021	All other reportable incidents	47263	Cardiac resynchronization therapy implantable pacemaker	CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	23/07/2021	23/07/2021
							F08	Hospitalization or Prolonged Hospitalization	A24	No Apparent Device Problem	23/07/2021	23/07/2021
							F1905	Device Revision or Replacement	A24	No Apparent Device Problem	23/07/2021	23/07/2021
					CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	23/07/2021	23/07/2021
							F08	Hospitalization or Prolonged Hospitalization	A24	No Apparent Device Problem	23/07/2021	23/07/2021
							F1905	Device Revision or Replacement	A24	No Apparent Device Problem	23/07/2021	23/07/2021
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	23/07/2021	23/07/2021
							F08	Hospitalization or Prolonged Hospitalization	A24	No Apparent Device Problem	23/07/2021	23/07/2021
							F1905	Device Revision or Replacement	A24	No Apparent Device Problem	23/07/2021	23/07/2021
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	23/07/2021	23/07/2021
							F08	Hospitalization or Prolonged Hospitalization	A24	No Apparent Device Problem	23/07/2021	23/07/2021
							F1905	Device Revision or Replacement	A24	No Apparent Device Problem	23/07/2021	23/07/2021
6419	2022	All other reportable incidents	35852	Single-chamber implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		

6420	2021	All other reportable incidents	47267	Single-chamber implantable pacemaker, rate-responsive	CT141	Electrodes/Leads and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	28/06/2021	28/06/2021
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	28/06/2021	28/06/2021
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	28/06/2021	28/06/2021
					CT145	Cardiovascular devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	28/06/2021	28/06/2021
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	28/06/2021	28/06/2021
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	28/06/2021	28/06/2021
					CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	28/06/2021	28/06/2021
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	28/06/2021	28/06/2021
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	28/06/2021	28/06/2021
					CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	28/06/2021	28/06/2021
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	28/06/2021	28/06/2021
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	28/06/2021	28/06/2021
					CT233	Surgical	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	28/06/2021	28/06/2021
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	28/06/2021	28/06/2021
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	28/06/2021	28/06/2021
					CT2406	Implantable	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	28/06/2021	28/06/2021
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	28/06/2021	28/06/2021
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	28/06/2021	28/06/2021
					CT2407	Active implantable	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	28/06/2021	28/06/2021
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	28/06/2021	28/06/2021
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	28/06/2021	28/06/2021
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	28/06/2021	28/06/2021
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	28/06/2021	28/06/2021
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	28/06/2021	28/06/2021
					CT2727	Active implant devices/ancillaries	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	28/06/2021	28/06/2021
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	28/06/2021	28/06/2021
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	28/06/2021	28/06/2021
					CT2750	Implantable cardiac electrical stimulators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	28/06/2021	28/06/2021
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	28/06/2021	28/06/2021
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	28/06/2021	28/06/2021
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	28/06/2021	28/06/2021
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	28/06/2021	28/06/2021
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	28/06/2021	28/06/2021
					CT2985	Active	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	28/06/2021	28/06/2021
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	28/06/2021	28/06/2021
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	28/06/2021	28/06/2021
					CT3001	Implantable metal	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	28/06/2021	28/06/2021
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	28/06/2021	28/06/2021
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	28/06/2021	28/06/2021
					CT3002	Implantable synthetic polymer	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	28/06/2021	28/06/2021
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	28/06/2021	28/06/2021
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	28/06/2021	28/06/2021
					CT3006	Implantable inorganic material	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	28/06/2021	28/06/2021
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	28/06/2021	28/06/2021
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	28/06/2021	28/06/2021
					CT314	Electrically powered	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	28/06/2021	28/06/2021
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	28/06/2021	28/06/2021
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	28/06/2021	28/06/2021
					CT321	Long-term surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	28/06/2021	28/06/2021
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	28/06/2021	28/06/2021
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	28/06/2021	28/06/2021
					CT449	Generators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	28/06/2021	28/06/2021
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	28/06/2021	28/06/2021
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	28/06/2021	28/06/2021
					CT981	Single use	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	28/06/2021	28/06/2021
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	28/06/2021	28/06/2021
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	28/06/2021	28/06/2021
					CT983	Surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	28/06/2021	28/06/2021
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	28/06/2021	28/06/2021
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	28/06/2021	28/06/2021
6421	2021	All other reportable incidents	35852	Single-chamber implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem	27/01/2022	27/01/2022
									A070504	Premature Discharge of Battery	27/01/2022	27/01/2022
							F08	Hospitalization or Prolonged Hospitalization	A0705	Battery Problem	27/01/2022	27/01/2022
									A070504	Premature Discharge of Battery	27/01/2022	27/01/2022
							F1905	Device Revision or Replacement	A0705	Battery Problem	27/01/2022	27/01/2022
									A070504	Premature Discharge of Battery	27/01/2022	27/01/2022
6422	2021	All other reportable incidents	47265	Dual-chamber implantable pacemaker, rate-responsive	CT141	Electrodes/Leads and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT145	Cardiovascular devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT233	Surgical	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT2406	Implantable	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT2407	Active implantable	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT2727	Active implant devices/ancillaries	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT2750	Implantable cardiac electrical stimulators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		

					CT2985	Active	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT3001	Implantable metal	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT3002	Implantable synthetic polymer	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT3006	Implantable inorganic material	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT314	Electrically powered	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT321	Long-term surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT449	Generators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT981	Single use	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT983	Surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
6423	2022	All other reportable incidents	37265	Dual-chamber implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E2402	Appropriate Clinical Signs, Symptoms and Conditions Term/Code not Available	A0709	Device Sensing Problem		
									A071301	Failure to Deliver Shock/Stimulation		
							E2403	No Clinical Signs, Symptoms or Conditions	A0709	Device Sensing Problem		
									A071301	Failure to Deliver Shock/Stimulation		
							F1001	Absence of Treatment	A0709	Device Sensing Problem		
									A071301	Failure to Deliver Shock/Stimulation		
							F1203	Life Threatening Illness or Injury	A0709	Device Sensing Problem		
									A071301	Failure to Deliver Shock/Stimulation		
6424	2022	All other reportable incidents	47267	Single-chamber implantable pacemaker, rate-responsive	CT141	Electrodes/Leads and associated devices	E0112	Dizziness	A070101	Failure to Capture		
									A072201	High Impedance		
							E2403	No Clinical Signs, Symptoms or Conditions	A070101	Failure to Capture		
									A072201	High Impedance		
							F26	No Health Consequences or Impact	A070101	Failure to Capture		
									A072201	High Impedance		
					CT145	Cardiovascular devices	E0112	Dizziness	A070101	Failure to Capture		
									A072201	High Impedance		
							E2403	No Clinical Signs, Symptoms or Conditions	A070101	Failure to Capture		
									A072201	High Impedance		
							F26	No Health Consequences or Impact	A070101	Failure to Capture		
									A072201	High Impedance		
					CT1684	Cardiac pacemakers and associated devices	E0112	Dizziness	A070101	Failure to Capture		
									A072201	High Impedance		
							E2403	No Clinical Signs, Symptoms or Conditions	A070101	Failure to Capture		
									A072201	High Impedance		
							F26	No Health Consequences or Impact	A070101	Failure to Capture		
									A072201	High Impedance		
					CT1685	Cardiac pacemakers	E0112	Dizziness	A070101	Failure to Capture		
									A072201	High Impedance		
							E2403	No Clinical Signs, Symptoms or Conditions	A070101	Failure to Capture		
									A072201	High Impedance		
							F26	No Health Consequences or Impact	A070101	Failure to Capture		
									A072201	High Impedance		
					CT233	Surgical	E0112	Dizziness	A070101	Failure to Capture		
									A072201	High Impedance		
							E2403	No Clinical Signs, Symptoms or Conditions	A070101	Failure to Capture		
									A072201	High Impedance		
							F26	No Health Consequences or Impact	A070101	Failure to Capture		
									A072201	High Impedance		
					CT2406	Implantable	E0112	Dizziness	A070101	Failure to Capture		
									A072201	High Impedance		
							E2403	No Clinical Signs, Symptoms or Conditions	A070101	Failure to Capture		
									A072201	High Impedance		
							F26	No Health Consequences or Impact	A070101	Failure to Capture		
									A072201	High Impedance		
					CT2407	Active implantable	E0112	Dizziness	A070101	Failure to Capture		
									A072201	High Impedance		
							E2403	No Clinical Signs, Symptoms or Conditions	A070101	Failure to Capture		
									A072201	High Impedance		
							F26	No Health Consequences or Impact	A070101	Failure to Capture		
									A072201	High Impedance		
					CT257	Implantable cardiac pacemakers	E0112	Dizziness	A070101	Failure to Capture		
									A072201	High Impedance		
							E2403	No Clinical Signs, Symptoms or Conditions	A070101	Failure to Capture		
									A072201	High Impedance		
							F26	No Health Consequences or Impact	A070101	Failure to Capture		
									A072201	High Impedance		
					CT2727	Active implant devices/ancillaries	E0112	Dizziness	A070101	Failure to Capture		
									A072201	High Impedance		
							E2403	No Clinical Signs, Symptoms or Conditions	A070101	Failure to Capture		
									A072201	High Impedance		
							F26	No Health Consequences or Impact	A070101	Failure to Capture		
									A072201	High Impedance		
					CT2750	Implantable cardiac electrical stimulators and associated devices	E0112	Dizziness	A070101	Failure to Capture		
									A072201	High Impedance		
							E2403	No Clinical Signs, Symptoms or Conditions	A070101	Failure to Capture		
									A072201	High Impedance		
							F26	No Health Consequences or Impact	A070101	Failure to Capture		
									A072201	High Impedance		
					CT2866	Implantable cardiac electrical stimulators	E0112	Dizziness	A070101	Failure to Capture		
									A072201	High Impedance		

							E2403	No Clinical Signs, Symptoms or Conditions	A070101	Failure to Capture		
									A072201	High Impedance		
							F26	No Health Consequences or Impact	A070101	Failure to Capture		
									A072201	High Impedance		
					CT2985	Active	E0112	Dizziness	A070101	Failure to Capture		
									A072201	High Impedance		
							E2403	No Clinical Signs, Symptoms or Conditions	A070101	Failure to Capture		
									A072201	High Impedance		
							F26	No Health Consequences or Impact	A070101	Failure to Capture		
									A072201	High Impedance		
					CT3001	Implantable metal	E0112	Dizziness	A070101	Failure to Capture		
									A072201	High Impedance		
							E2403	No Clinical Signs, Symptoms or Conditions	A070101	Failure to Capture		
									A072201	High Impedance		
							F26	No Health Consequences or Impact	A070101	Failure to Capture		
									A072201	High Impedance		
					CT3002	Implantable synthetic polymer	E0112	Dizziness	A070101	Failure to Capture		
									A072201	High Impedance		
							E2403	No Clinical Signs, Symptoms or Conditions	A070101	Failure to Capture		
									A072201	High Impedance		
							F26	No Health Consequences or Impact	A070101	Failure to Capture		
									A072201	High Impedance		
					CT3006	Implantable inorganic material	E0112	Dizziness	A070101	Failure to Capture		
									A072201	High Impedance		
							E2403	No Clinical Signs, Symptoms or Conditions	A070101	Failure to Capture		
									A072201	High Impedance		
							F26	No Health Consequences or Impact	A070101	Failure to Capture		
									A072201	High Impedance		
					CT314	Electrically powered	E0112	Dizziness	A070101	Failure to Capture		
									A072201	High Impedance		
							E2403	No Clinical Signs, Symptoms or Conditions	A070101	Failure to Capture		
									A072201	High Impedance		
							F26	No Health Consequences or Impact	A070101	Failure to Capture		
									A072201	High Impedance		
					CT321	Long-term surgical invasive	E0112	Dizziness	A070101	Failure to Capture		
									A072201	High Impedance		
							E2403	No Clinical Signs, Symptoms or Conditions	A070101	Failure to Capture		
									A072201	High Impedance		
							F26	No Health Consequences or Impact	A070101	Failure to Capture		
									A072201	High Impedance		
					CT449	Generators and associated devices	E0112	Dizziness	A070101	Failure to Capture		
									A072201	High Impedance		
							E2403	No Clinical Signs, Symptoms or Conditions	A070101	Failure to Capture		
									A072201	High Impedance		
							F26	No Health Consequences or Impact	A070101	Failure to Capture		
									A072201	High Impedance		
					CT981	Single use	E0112	Dizziness	A070101	Failure to Capture		
									A072201	High Impedance		
							E2403	No Clinical Signs, Symptoms or Conditions	A070101	Failure to Capture		
									A072201	High Impedance		
							F26	No Health Consequences or Impact	A070101	Failure to Capture		
									A072201	High Impedance		
					CT983	Surgical invasive	E0112	Dizziness	A070101	Failure to Capture		
									A072201	High Impedance		
							E2403	No Clinical Signs, Symptoms or Conditions	A070101	Failure to Capture		
									A072201	High Impedance		
							F26	No Health Consequences or Impact	A070101	Failure to Capture		
									A072201	High Impedance		
6425	2022	All other reportable incidents	17882	Physiologic-monitoring defibrillation system	CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A12	Connection Problem		
							F26	No Health Consequences or Impact	A12	Connection Problem		
6426	2022	All other reportable incidents	17882	Physiologic-monitoring defibrillation system	CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070803	Failure to Power Up		
							F27	Problem Identified During Non-Clinical Procedure	A070803	Failure to Power Up		
6427	2022	All other reportable incidents	47265	Dual-chamber implantable pacemaker, rate-responsive	CT141	Electrodes/Leads and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	06/05/2022	06/05/2022
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	06/05/2022	06/05/2022
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	06/05/2022	06/05/2022
					CT145	Cardiovascular devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	06/05/2022	06/05/2022
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	06/05/2022	06/05/2022
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	06/05/2022	06/05/2022
					CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	06/05/2022	06/05/2022
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	06/05/2022	06/05/2022
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	06/05/2022	06/05/2022
					CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	06/05/2022	06/05/2022
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	06/05/2022	06/05/2022
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	06/05/2022	06/05/2022
					CT233	Surgical	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	06/05/2022	06/05/2022
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	06/05/2022	06/05/2022
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	06/05/2022	06/05/2022
					CT2406	Implantable	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	06/05/2022	06/05/2022
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	06/05/2022	06/05/2022
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	06/05/2022	06/05/2022
					CT2407	Active implantable	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	06/05/2022	06/05/2022
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	06/05/2022	06/05/2022
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	06/05/2022	06/05/2022
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	06/05/2022	06/05/2022
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	06/05/2022	06/05/2022
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	06/05/2022	06/05/2022
					CT2727	Active implant devices/ancillaries	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	06/05/2022	06/05/2022
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	06/05/2022	06/05/2022

							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	06/05/2022	06/05/2022
					CT2750	Implantable cardiac electrical stimulators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	06/05/2022	06/05/2022
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	06/05/2022	06/05/2022
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	06/05/2022	06/05/2022
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	06/05/2022	06/05/2022
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	06/05/2022	06/05/2022
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	06/05/2022	06/05/2022
					CT2985	Active	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	06/05/2022	06/05/2022
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	06/05/2022	06/05/2022
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	06/05/2022	06/05/2022
					CT3001	Implantable metal	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	06/05/2022	06/05/2022
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	06/05/2022	06/05/2022
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	06/05/2022	06/05/2022
					CT3002	Implantable synthetic polymer	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	06/05/2022	06/05/2022
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	06/05/2022	06/05/2022
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	06/05/2022	06/05/2022
					CT3006	Implantable inorganic material	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	06/05/2022	06/05/2022
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	06/05/2022	06/05/2022
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	06/05/2022	06/05/2022
					CT314	Electrically powered	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	06/05/2022	06/05/2022
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	06/05/2022	06/05/2022
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	06/05/2022	06/05/2022
					CT321	Long-term surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	06/05/2022	06/05/2022
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	06/05/2022	06/05/2022
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	06/05/2022	06/05/2022
					CT449	Generators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	06/05/2022	06/05/2022
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	06/05/2022	06/05/2022
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	06/05/2022	06/05/2022
					CT981	Single use	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	06/05/2022	06/05/2022
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	06/05/2022	06/05/2022
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	06/05/2022	06/05/2022
					CT983	Surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	06/05/2022	06/05/2022
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	06/05/2022	06/05/2022
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	06/05/2022	06/05/2022
6428	2021	All other reportable incidents	47265	Dual-chamber implantable pacemaker, rate-responsive	CT141	Electrodes/Leads and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070101	Failure to Capture		
									A070102	High Capture Threshold		
							F26	No Health Consequences or Impact	A070101	Failure to Capture		
									A070102	High Capture Threshold		
					CT145	Cardiovascular devices	E2403	No Clinical Signs, Symptoms or Conditions	A070101	Failure to Capture		
									A070102	High Capture Threshold		
							F26	No Health Consequences or Impact	A070101	Failure to Capture		
									A070102	High Capture Threshold		
					CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070101	Failure to Capture		
									A070102	High Capture Threshold		
							F26	No Health Consequences or Impact	A070101	Failure to Capture		
									A070102	High Capture Threshold		
					CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A070101	Failure to Capture		
									A070102	High Capture Threshold		
							F26	No Health Consequences or Impact	A070101	Failure to Capture		
									A070102	High Capture Threshold		
					CT233	Surgical	E2403	No Clinical Signs, Symptoms or Conditions	A070101	Failure to Capture		
									A070102	High Capture Threshold		
							F26	No Health Consequences or Impact	A070101	Failure to Capture		
									A070102	High Capture Threshold		
					CT2406	Implantable	E2403	No Clinical Signs, Symptoms or Conditions	A070101	Failure to Capture		
									A070102	High Capture Threshold		
							F26	No Health Consequences or Impact	A070101	Failure to Capture		
									A070102	High Capture Threshold		
					CT2407	Active implantable	E2403	No Clinical Signs, Symptoms or Conditions	A070101	Failure to Capture		
									A070102	High Capture Threshold		
							F26	No Health Consequences or Impact	A070101	Failure to Capture		
									A070102	High Capture Threshold		
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A070101	Failure to Capture		
									A070102	High Capture Threshold		
							F26	No Health Consequences or Impact	A070101	Failure to Capture		
									A070102	High Capture Threshold		
					CT2727	Active implant devices/ancillaries	E2403	No Clinical Signs, Symptoms or Conditions	A070101	Failure to Capture		
									A070102	High Capture Threshold		
							F26	No Health Consequences or Impact	A070101	Failure to Capture		
									A070102	High Capture Threshold		
					CT2750	Implantable cardiac electrical stimulators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070101	Failure to Capture		
									A070102	High Capture Threshold		
							F26	No Health Consequences or Impact	A070101	Failure to Capture		
									A070102	High Capture Threshold		
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A070101	Failure to Capture		
									A070102	High Capture Threshold		
							F26	No Health Consequences or Impact	A070101	Failure to Capture		
									A070102	High Capture Threshold		
					CT2985	Active	E2403	No Clinical Signs, Symptoms or Conditions	A070101	Failure to Capture		
									A070102	High Capture Threshold		
							F26	No Health Consequences or Impact	A070101	Failure to Capture		
									A070102	High Capture Threshold		
					CT3001	Implantable metal	E2403	No Clinical Signs, Symptoms or Conditions	A070101	Failure to Capture		
									A070102	High Capture Threshold		
							F26	No Health Consequences or Impact	A070101	Failure to Capture		
									A070102	High Capture Threshold		
					CT3002	Implantable synthetic polymer	E2403	No Clinical Signs, Symptoms or Conditions	A070101	Failure to Capture		
									A070102	High Capture Threshold		

						F26	No Health Consequences or Impact	A070101	Failure to Capture		
								A070102	High Capture Threshold		
					CT3006	Implantable inorganic material	E2403	No Clinical Signs, Symptoms or Conditions	A070101	Failure to Capture		
									A070102	High Capture Threshold		
							F26	No Health Consequences or Impact	A070101	Failure to Capture		
									A070102	High Capture Threshold		
					CT314	Electrically powered	E2403	No Clinical Signs, Symptoms or Conditions	A070101	Failure to Capture		
									A070102	High Capture Threshold		
							F26	No Health Consequences or Impact	A070101	Failure to Capture		
									A070102	High Capture Threshold		
					CT321	Long-term surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A070101	Failure to Capture		
									A070102	High Capture Threshold		
							F26	No Health Consequences or Impact	A070101	Failure to Capture		
									A070102	High Capture Threshold		
					CT449	Generators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070101	Failure to Capture		
									A070102	High Capture Threshold		
							F26	No Health Consequences or Impact	A070101	Failure to Capture		
									A070102	High Capture Threshold		
					CT981	Single use	E2403	No Clinical Signs, Symptoms or Conditions	A070101	Failure to Capture		
									A070102	High Capture Threshold		
							F26	No Health Consequences or Impact	A070101	Failure to Capture		
									A070102	High Capture Threshold		
					CT983	Surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A070101	Failure to Capture		
									A070102	High Capture Threshold		
							F26	No Health Consequences or Impact	A070101	Failure to Capture		
									A070102	High Capture Threshold		
6429	2022	All other reportable incidents	17882	Physiologic-monitoring defibrillation system	CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070803	Failure to Power Up		
							F27	Problem Identified During Non-Clinical Procedure	A070803	Failure to Power Up		
6430	2021	All other reportable incidents	47265	Dual-chamber implantable pacemaker, rate-responsive	CT141	Electrodes/Leads and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070909	Oversensing	24/08/2021	24/08/2021
									A1205	Loose or Intermittent Connection	24/08/2021	24/08/2021
							F1901	Additional Surgery	A070909	Oversensing	24/08/2021	24/08/2021
									A1205	Loose or Intermittent Connection	24/08/2021	24/08/2021
					CT145	Cardiovascular devices	E2403	No Clinical Signs, Symptoms or Conditions	A070909	Oversensing	24/08/2021	24/08/2021
									A1205	Loose or Intermittent Connection	24/08/2021	24/08/2021
							F1901	Additional Surgery	A070909	Oversensing	24/08/2021	24/08/2021
									A1205	Loose or Intermittent Connection	24/08/2021	24/08/2021
					CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070909	Oversensing	24/08/2021	24/08/2021
									A1205	Loose or Intermittent Connection	24/08/2021	24/08/2021
							F1901	Additional Surgery	A070909	Oversensing	24/08/2021	24/08/2021
									A1205	Loose or Intermittent Connection	24/08/2021	24/08/2021
					CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A070909	Oversensing	24/08/2021	24/08/2021
									A1205	Loose or Intermittent Connection	24/08/2021	24/08/2021
							F1901	Additional Surgery	A070909	Oversensing	24/08/2021	24/08/2021
									A1205	Loose or Intermittent Connection	24/08/2021	24/08/2021
					CT233	Surgical	E2403	No Clinical Signs, Symptoms or Conditions	A070909	Oversensing	24/08/2021	24/08/2021
									A1205	Loose or Intermittent Connection	24/08/2021	24/08/2021
							F1901	Additional Surgery	A070909	Oversensing	24/08/2021	24/08/2021
									A1205	Loose or Intermittent Connection	24/08/2021	24/08/2021
					CT2406	Implantable	E2403	No Clinical Signs, Symptoms or Conditions	A070909	Oversensing	24/08/2021	24/08/2021
									A1205	Loose or Intermittent Connection	24/08/2021	24/08/2021
							F1901	Additional Surgery	A070909	Oversensing	24/08/2021	24/08/2021
									A1205	Loose or Intermittent Connection	24/08/2021	24/08/2021
					CT2407	Active implantable	E2403	No Clinical Signs, Symptoms or Conditions	A070909	Oversensing	24/08/2021	24/08/2021
									A1205	Loose or Intermittent Connection	24/08/2021	24/08/2021
							F1901	Additional Surgery	A070909	Oversensing	24/08/2021	24/08/2021
									A1205	Loose or Intermittent Connection	24/08/2021	24/08/2021
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A070909	Oversensing	24/08/2021	24/08/2021
									A1205	Loose or Intermittent Connection	24/08/2021	24/08/2021
							F1901	Additional Surgery	A070909	Oversensing	24/08/2021	24/08/2021
									A1205	Loose or Intermittent Connection	24/08/2021	24/08/2021
					CT2727	Active implant devices/ancillaries	E2403	No Clinical Signs, Symptoms or Conditions	A070909	Oversensing	24/08/2021	24/08/2021
									A1205	Loose or Intermittent Connection	24/08/2021	24/08/2021
							F1901	Additional Surgery	A070909	Oversensing	24/08/2021	24/08/2021
									A1205	Loose or Intermittent Connection	24/08/2021	24/08/2021
					CT2750	Implantable cardiac electrical stimulators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070909	Oversensing	24/08/2021	24/08/2021
									A1205	Loose or Intermittent Connection	24/08/2021	24/08/2021
							F1901	Additional Surgery	A070909	Oversensing	24/08/2021	24/08/2021
									A1205	Loose or Intermittent Connection	24/08/2021	24/08/2021
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A070909	Oversensing	24/08/2021	24/08/2021
									A1205	Loose or Intermittent Connection	24/08/2021	24/08/2021
							F1901	Additional Surgery	A070909	Oversensing	24/08/2021	24/08/2021
									A1205	Loose or Intermittent Connection	24/08/2021	24/08/2021
					CT2985	Active	E2403	No Clinical Signs, Symptoms or Conditions	A070909	Oversensing	24/08/2021	24/08/2021
									A1205	Loose or Intermittent Connection	24/08/2021	24/08/2021
							F1901	Additional Surgery	A070909	Oversensing	24/08/2021	24/08/2021
									A1205	Loose or Intermittent Connection	24/08/2021	24/08/2021
					CT3001	Implantable metal	E2403	No Clinical Signs, Symptoms or Conditions	A070909	Oversensing	24/08/2021	24/08/2021
									A1205	Loose or Intermittent Connection	24/08/2021	24/08/2021
							F1901	Additional Surgery	A070909	Oversensing	24/08/2021	24/08/2021
									A1205	Loose or Intermittent Connection	24/08/2021	24/08/2021
					CT3002	Implantable synthetic polymer	E2403	No Clinical Signs, Symptoms or Conditions	A070909	Oversensing	24/08/2021	24/08/2021
									A1205	Loose or Intermittent Connection	24/08/2021	24/08/2021
							F1901	Additional Surgery	A070909	Oversensing	24/08/2021	24/08/2021
									A1205	Loose or Intermittent Connection	24/08/2021	24/08/2021
					CT3006	Implantable inorganic material	E2403	No Clinical Signs, Symptoms or Conditions	A070909	Oversensing	24/08/2021	24/08/2021
									A1205	Loose or Intermittent Connection	24/08/2021	24/08/2021
							F1901	Additional Surgery	A070909	Oversensing	24/08/2021	24/08/2021
									A1205	Loose or Intermittent Connection	24/08/2021	24/08/2021

					CT314	Electrically powered	E2403	No Clinical Signs, Symptoms or Conditions	A070909	Oversensing	24/08/2021	24/08/2021
									A1205	Loose or Intermittent Connection	24/08/2021	24/08/2021
							F1901	Additional Surgery	A070909	Oversensing	24/08/2021	24/08/2021
									A1205	Loose or Intermittent Connection	24/08/2021	24/08/2021
					CT321	Long-term surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A070909	Oversensing	24/08/2021	24/08/2021
									A1205	Loose or Intermittent Connection	24/08/2021	24/08/2021
							F1901	Additional Surgery	A070909	Oversensing	24/08/2021	24/08/2021
									A1205	Loose or Intermittent Connection	24/08/2021	24/08/2021
					CT449	Generators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070909	Oversensing	24/08/2021	24/08/2021
									A1205	Loose or Intermittent Connection	24/08/2021	24/08/2021
							F1901	Additional Surgery	A070909	Oversensing	24/08/2021	24/08/2021
									A1205	Loose or Intermittent Connection	24/08/2021	24/08/2021
					CT981	Single use	E2403	No Clinical Signs, Symptoms or Conditions	A070909	Oversensing	24/08/2021	24/08/2021
									A1205	Loose or Intermittent Connection	24/08/2021	24/08/2021
							F1901	Additional Surgery	A070909	Oversensing	24/08/2021	24/08/2021
									A1205	Loose or Intermittent Connection	24/08/2021	24/08/2021
					CT983	Surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A070909	Oversensing	24/08/2021	24/08/2021
									A1205	Loose or Intermittent Connection	24/08/2021	24/08/2021
							F1901	Additional Surgery	A070909	Oversensing	24/08/2021	24/08/2021
									A1205	Loose or Intermittent Connection	24/08/2021	24/08/2021
6431	2022	All other reportable incidents	35223	Endocardial/interventricular septal pacing lead	CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A072202	Low Impedance	14/04/2022	14/04/2022
							F1908	Prolonged Surgery	A072202	Low Impedance	14/04/2022	14/04/2022
6432	2021	All other reportable incidents	17882	Physiologic-monitoring defibrillation system	CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A07	Electrical /Electronic Property Problem		
							F27	Problem Identified During Non-Clinical Procedure	A07	Electrical /Electronic Property Problem		
6433	2021	All other reportable incidents	17882	Physiologic-monitoring defibrillation system	CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070802	Intermittent Loss of Power		
							F27	Problem Identified During Non-Clinical Procedure	A070802	Intermittent Loss of Power		
6434	2021	All other reportable incidents	63188	Left-ventricular wireless pacing system sensor/transmitter	CT141	Electrodes/Leads and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A0710	Failure to Conduct		
					CT145	Cardiovascular devices	E2403	No Clinical Signs, Symptoms or Conditions	A0710	Failure to Conduct		
					CT1577	Wireless monitoring system transmitters	E2403	No Clinical Signs, Symptoms or Conditions	A0710	Failure to Conduct		
					CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A0710	Failure to Conduct		
					CT2317	Transmitters/Receivers/Transceivers	E2403	No Clinical Signs, Symptoms or Conditions	A0710	Failure to Conduct		
					CT233	Surgical	E2403	No Clinical Signs, Symptoms or Conditions	A0710	Failure to Conduct		
					CT2406	Implantable	E2403	No Clinical Signs, Symptoms or Conditions	A0710	Failure to Conduct		
					CT2407	Active implantable	E2403	No Clinical Signs, Symptoms or Conditions	A0710	Failure to Conduct		
					CT255	Transducers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A0710	Failure to Conduct		
					CT2726	Physiological monitoring/measuring devices	E2403	No Clinical Signs, Symptoms or Conditions	A0710	Failure to Conduct		
					CT2727	Active implant devices/ancillaries	E2403	No Clinical Signs, Symptoms or Conditions	A0710	Failure to Conduct		
					CT2750	Implantable cardiac electrical stimulators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A0710	Failure to Conduct		
					CT2885	Physiological monitoring/measuring system transmitters/receivers/transceivers	E2403	No Clinical Signs, Symptoms or Conditions	A0710	Failure to Conduct		
					CT2985	Active	E2403	No Clinical Signs, Symptoms or Conditions	A0710	Failure to Conduct		
					CT3001	Implantable metal	E2403	No Clinical Signs, Symptoms or Conditions	A0710	Failure to Conduct		
					CT3002	Implantable synthetic polymer	E2403	No Clinical Signs, Symptoms or Conditions	A0710	Failure to Conduct		
					CT3006	Implantable inorganic material	E2403	No Clinical Signs, Symptoms or Conditions	A0710	Failure to Conduct		
					CT314	Electrically powered	E2403	No Clinical Signs, Symptoms or Conditions	A0710	Failure to Conduct		
					CT321	Long-term surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A0710	Failure to Conduct		
					CT384	Sensors and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A0710	Failure to Conduct		
					CT449	Generators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A0710	Failure to Conduct		
					CT981	Single use	E2403	No Clinical Signs, Symptoms or Conditions	A0710	Failure to Conduct		
					CT983	Surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A0710	Failure to Conduct		
					CT994	Sensors	E2403	No Clinical Signs, Symptoms or Conditions	A0710	Failure to Conduct		
6435	2022	All other reportable incidents	47267	Single-chamber implantable pacemaker, rate-responsive	CT141	Electrodes/Leads and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A0701	Capturing Problem		
									A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A0701	Capturing Problem		
									A070504	Premature Discharge of Battery		
					CT145	Cardiovascular devices	E2403	No Clinical Signs, Symptoms or Conditions	A0701	Capturing Problem		
									A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A0701	Capturing Problem		
									A070504	Premature Discharge of Battery		
					CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A0701	Capturing Problem		
									A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A0701	Capturing Problem		
									A070504	Premature Discharge of Battery		
					CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A0701	Capturing Problem		
									A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A0701	Capturing Problem		
									A070504	Premature Discharge of Battery		
					CT233	Surgical	E2403	No Clinical Signs, Symptoms or Conditions	A0701	Capturing Problem		
									A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A0701	Capturing Problem		
									A070504	Premature Discharge of Battery		
					CT2406	Implantable	E2403	No Clinical Signs, Symptoms or Conditions	A0701	Capturing Problem		
									A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A0701	Capturing Problem		
									A070504	Premature Discharge of Battery		
					CT2407	Active implantable	E2403	No Clinical Signs, Symptoms or Conditions	A0701	Capturing Problem		
									A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A0701	Capturing Problem		
									A070504	Premature Discharge of Battery		
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A0701	Capturing Problem		
									A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A0701	Capturing Problem		
									A070504	Premature Discharge of Battery		
					CT2727	Active implant devices/ancillaries	E2403	No Clinical Signs, Symptoms or Conditions	A0701	Capturing Problem		
									A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A0701	Capturing Problem		
									A070504	Premature Discharge of Battery		

					CT2750	Implantable cardiac electrical stimulators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A0701	Capturing Problem		
									A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A0701	Capturing Problem		
									A070504	Premature Discharge of Battery		
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A0701	Capturing Problem		
									A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A0701	Capturing Problem		
									A070504	Premature Discharge of Battery		
					CT2985	Active	E2403	No Clinical Signs, Symptoms or Conditions	A0701	Capturing Problem		
									A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A0701	Capturing Problem		
									A070504	Premature Discharge of Battery		
					CT3001	Implantable metal	E2403	No Clinical Signs, Symptoms or Conditions	A0701	Capturing Problem		
									A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A0701	Capturing Problem		
									A070504	Premature Discharge of Battery		
					CT3002	Implantable synthetic polymer	E2403	No Clinical Signs, Symptoms or Conditions	A0701	Capturing Problem		
									A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A0701	Capturing Problem		
									A070504	Premature Discharge of Battery		
					CT3006	Implantable inorganic material	E2403	No Clinical Signs, Symptoms or Conditions	A0701	Capturing Problem		
									A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A0701	Capturing Problem		
									A070504	Premature Discharge of Battery		
					CT314	Electrically powered	E2403	No Clinical Signs, Symptoms or Conditions	A0701	Capturing Problem		
									A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A0701	Capturing Problem		
									A070504	Premature Discharge of Battery		
					CT321	Long-term surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A0701	Capturing Problem		
									A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A0701	Capturing Problem		
									A070504	Premature Discharge of Battery		
					CT449	Generators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A0701	Capturing Problem		
									A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A0701	Capturing Problem		
									A070504	Premature Discharge of Battery		
					CT981	Single use	E2403	No Clinical Signs, Symptoms or Conditions	A0701	Capturing Problem		
									A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A0701	Capturing Problem		
									A070504	Premature Discharge of Battery		
					CT983	Surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A0701	Capturing Problem		
									A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A0701	Capturing Problem		
									A070504	Premature Discharge of Battery		
6436	2022	All other reportable incidents	47270	Cardiac resynchronization therapy implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E0112	Dizziness	A070909	Oversensing		
							E2403	No Clinical Signs, Symptoms or Conditions	A070909	Oversensing		
							F23	Unexpected Medical Intervention	A070909	Oversensing		
6437	2022	All other reportable incidents	47265	Dual-chamber implantable pacemaker, rate-responsive	CT141	Electrodes/Leads and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	13/04/2022	13/04/2022
							F1901	Additional Surgery	A070504	Premature Discharge of Battery	13/04/2022	13/04/2022
					CT145	Cardiovascular devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	13/04/2022	13/04/2022
							F1901	Additional Surgery	A070504	Premature Discharge of Battery	13/04/2022	13/04/2022
					CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	13/04/2022	13/04/2022
							F1901	Additional Surgery	A070504	Premature Discharge of Battery	13/04/2022	13/04/2022
					CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	13/04/2022	13/04/2022
							F1901	Additional Surgery	A070504	Premature Discharge of Battery	13/04/2022	13/04/2022
					CT233	Surgical	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	13/04/2022	13/04/2022
							F1901	Additional Surgery	A070504	Premature Discharge of Battery	13/04/2022	13/04/2022
					CT2406	Implantable	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	13/04/2022	13/04/2022
							F1901	Additional Surgery	A070504	Premature Discharge of Battery	13/04/2022	13/04/2022
					CT2407	Active implantable	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	13/04/2022	13/04/2022
							F1901	Additional Surgery	A070504	Premature Discharge of Battery	13/04/2022	13/04/2022
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	13/04/2022	13/04/2022
							F1901	Additional Surgery	A070504	Premature Discharge of Battery	13/04/2022	13/04/2022
					CT2727	Active implant devices/ancillaries	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	13/04/2022	13/04/2022
							F1901	Additional Surgery	A070504	Premature Discharge of Battery	13/04/2022	13/04/2022
					CT2750	Implantable cardiac electrical stimulators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	13/04/2022	13/04/2022
							F1901	Additional Surgery	A070504	Premature Discharge of Battery	13/04/2022	13/04/2022
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	13/04/2022	13/04/2022
							F1901	Additional Surgery	A070504	Premature Discharge of Battery	13/04/2022	13/04/2022
					CT2985	Active	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	13/04/2022	13/04/2022
							F1901	Additional Surgery	A070504	Premature Discharge of Battery	13/04/2022	13/04/2022
					CT3001	Implantable metal	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	13/04/2022	13/04/2022
							F1901	Additional Surgery	A070504	Premature Discharge of Battery	13/04/2022	13/04/2022
					CT3002	Implantable synthetic polymer	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	13/04/2022	13/04/2022
							F1901	Additional Surgery	A070504	Premature Discharge of Battery	13/04/2022	13/04/2022
					CT3006	Implantable inorganic material	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	13/04/2022	13/04/2022
							F1901	Additional Surgery	A070504	Premature Discharge of Battery	13/04/2022	13/04/2022
					CT314	Electrically powered	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	13/04/2022	13/04/2022
							F1901	Additional Surgery	A070504	Premature Discharge of Battery	13/04/2022	13/04/2022
					CT321	Long-term surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	13/04/2022	13/04/2022
							F1901	Additional Surgery	A070504	Premature Discharge of Battery	13/04/2022	13/04/2022
					CT449	Generators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	13/04/2022	13/04/2022
							F1901	Additional Surgery	A070504	Premature Discharge of Battery	13/04/2022	13/04/2022
					CT981	Single use	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	13/04/2022	13/04/2022
							F1901	Additional Surgery	A070504	Premature Discharge of Battery	13/04/2022	13/04/2022
					CT983	Surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	13/04/2022	13/04/2022
							F1901	Additional Surgery	A070504	Premature Discharge of Battery	13/04/2022	13/04/2022
6438	2022	All other reportable incidents	47265	Dual-chamber implantable pacemaker, rate-	CT141	Electrodes/Leads and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem	13/06/2022	13/06/2022

				responsive					A070504	Premature Discharge of Battery	13/06/2022	13/06/2022
							F08	Hospitalization or Prolonged Hospitalization	A0705	Battery Problem	13/06/2022	13/06/2022
									A070504	Premature Discharge of Battery	13/06/2022	13/06/2022
							F1905	Device Revision or Replacement	A0705	Battery Problem	13/06/2022	13/06/2022
									A070504	Premature Discharge of Battery	13/06/2022	13/06/2022
					CT145	Cardiovascular devices	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem	13/06/2022	13/06/2022
									A070504	Premature Discharge of Battery	13/06/2022	13/06/2022
							F08	Hospitalization or Prolonged Hospitalization	A0705	Battery Problem	13/06/2022	13/06/2022
									A070504	Premature Discharge of Battery	13/06/2022	13/06/2022
							F1905	Device Revision or Replacement	A0705	Battery Problem	13/06/2022	13/06/2022
									A070504	Premature Discharge of Battery	13/06/2022	13/06/2022
					CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem	13/06/2022	13/06/2022
									A070504	Premature Discharge of Battery	13/06/2022	13/06/2022
							F08	Hospitalization or Prolonged Hospitalization	A0705	Battery Problem	13/06/2022	13/06/2022
									A070504	Premature Discharge of Battery	13/06/2022	13/06/2022
							F1905	Device Revision or Replacement	A0705	Battery Problem	13/06/2022	13/06/2022
									A070504	Premature Discharge of Battery	13/06/2022	13/06/2022
					CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem	13/06/2022	13/06/2022
									A070504	Premature Discharge of Battery	13/06/2022	13/06/2022
							F08	Hospitalization or Prolonged Hospitalization	A0705	Battery Problem	13/06/2022	13/06/2022
									A070504	Premature Discharge of Battery	13/06/2022	13/06/2022
							F1905	Device Revision or Replacement	A0705	Battery Problem	13/06/2022	13/06/2022
									A070504	Premature Discharge of Battery	13/06/2022	13/06/2022
					CT233	Surgical	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem	13/06/2022	13/06/2022
									A070504	Premature Discharge of Battery	13/06/2022	13/06/2022
							F08	Hospitalization or Prolonged Hospitalization	A0705	Battery Problem	13/06/2022	13/06/2022
									A070504	Premature Discharge of Battery	13/06/2022	13/06/2022
							F1905	Device Revision or Replacement	A0705	Battery Problem	13/06/2022	13/06/2022
									A070504	Premature Discharge of Battery	13/06/2022	13/06/2022
					CT2406	Implantable	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem	13/06/2022	13/06/2022
									A070504	Premature Discharge of Battery	13/06/2022	13/06/2022
							F08	Hospitalization or Prolonged Hospitalization	A0705	Battery Problem	13/06/2022	13/06/2022
									A070504	Premature Discharge of Battery	13/06/2022	13/06/2022
							F1905	Device Revision or Replacement	A0705	Battery Problem	13/06/2022	13/06/2022
									A070504	Premature Discharge of Battery	13/06/2022	13/06/2022
					CT2407	Active implantable	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem	13/06/2022	13/06/2022
									A070504	Premature Discharge of Battery	13/06/2022	13/06/2022
							F08	Hospitalization or Prolonged Hospitalization	A0705	Battery Problem	13/06/2022	13/06/2022
									A070504	Premature Discharge of Battery	13/06/2022	13/06/2022
							F1905	Device Revision or Replacement	A0705	Battery Problem	13/06/2022	13/06/2022
									A070504	Premature Discharge of Battery	13/06/2022	13/06/2022
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem	13/06/2022	13/06/2022
									A070504	Premature Discharge of Battery	13/06/2022	13/06/2022
							F08	Hospitalization or Prolonged Hospitalization	A0705	Battery Problem	13/06/2022	13/06/2022
									A070504	Premature Discharge of Battery	13/06/2022	13/06/2022
							F1905	Device Revision or Replacement	A0705	Battery Problem	13/06/2022	13/06/2022
									A070504	Premature Discharge of Battery	13/06/2022	13/06/2022
					CT2727	Active implant devices/ancillaries	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem	13/06/2022	13/06/2022
									A070504	Premature Discharge of Battery	13/06/2022	13/06/2022
							F08	Hospitalization or Prolonged Hospitalization	A0705	Battery Problem	13/06/2022	13/06/2022
									A070504	Premature Discharge of Battery	13/06/2022	13/06/2022
							F1905	Device Revision or Replacement	A0705	Battery Problem	13/06/2022	13/06/2022
									A070504	Premature Discharge of Battery	13/06/2022	13/06/2022
					CT2750	Implantable cardiac electrical stimulators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem	13/06/2022	13/06/2022
									A070504	Premature Discharge of Battery	13/06/2022	13/06/2022
							F08	Hospitalization or Prolonged Hospitalization	A0705	Battery Problem	13/06/2022	13/06/2022
									A070504	Premature Discharge of Battery	13/06/2022	13/06/2022
							F1905	Device Revision or Replacement	A0705	Battery Problem	13/06/2022	13/06/2022
									A070504	Premature Discharge of Battery	13/06/2022	13/06/2022
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem	13/06/2022	13/06/2022
									A070504	Premature Discharge of Battery	13/06/2022	13/06/2022
							F08	Hospitalization or Prolonged Hospitalization	A0705	Battery Problem	13/06/2022	13/06/2022
									A070504	Premature Discharge of Battery	13/06/2022	13/06/2022
							F1905	Device Revision or Replacement	A0705	Battery Problem	13/06/2022	13/06/2022
									A070504	Premature Discharge of Battery	13/06/2022	13/06/2022
					CT2985	Active	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem	13/06/2022	13/06/2022
									A070504	Premature Discharge of Battery	13/06/2022	13/06/2022
							F08	Hospitalization or Prolonged Hospitalization	A0705	Battery Problem	13/06/2022	13/06/2022
									A070504	Premature Discharge of Battery	13/06/2022	13/06/2022
							F1905	Device Revision or Replacement	A0705	Battery Problem	13/06/2022	13/06/2022
									A070504	Premature Discharge of Battery	13/06/2022	13/06/2022
					CT3001	Implantable metal	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem	13/06/2022	13/06/2022
									A070504	Premature Discharge of Battery	13/06/2022	13/06/2022
							F08	Hospitalization or Prolonged Hospitalization	A0705	Battery Problem	13/06/2022	13/06/2022
									A070504	Premature Discharge of Battery	13/06/2022	13/06/2022
							F1905	Device Revision or Replacement	A0705	Battery Problem	13/06/2022	13/06/2022
									A070504	Premature Discharge of Battery	13/06/2022	13/06/2022
					CT3002	Implantable synthetic polymer	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem	13/06/2022	13/06/2022
									A070504	Premature Discharge of Battery	13/06/2022	13/06/2022
							F08	Hospitalization or Prolonged Hospitalization	A0705	Battery Problem	13/06/2022	13/06/2022
									A070504	Premature Discharge of Battery	13/06/2022	13/06/2022
							F1905	Device Revision or Replacement	A0705	Battery Problem	13/06/2022	13/06/2022
									A070504	Premature Discharge of Battery	13/06/2022	13/06/2022
					CT3006	Implantable inorganic material	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem	13/06/2022	13/06/2022
									A070504	Premature Discharge of Battery	13/06/2022	13/06/2022
							F08	Hospitalization or Prolonged Hospitalization	A0705	Battery Problem	13/06/2022	13/06/2022
									A070504	Premature Discharge of Battery	13/06/2022	13/06/2022
							F1905	Device Revision or Replacement	A0705	Battery Problem	13/06/2022	13/06/2022

									A070504	Premature Discharge of Battery	13/06/2022	13/06/2022
					CT314	Electrically powered	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem	13/06/2022	13/06/2022
									A070504	Premature Discharge of Battery	13/06/2022	13/06/2022
							F08	Hospitalization or Prolonged Hospitalization	A0705	Battery Problem	13/06/2022	13/06/2022
									A070504	Premature Discharge of Battery	13/06/2022	13/06/2022
							F1905	Device Revision or Replacement	A0705	Battery Problem	13/06/2022	13/06/2022
									A070504	Premature Discharge of Battery	13/06/2022	13/06/2022
					CT321	Long-term surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem	13/06/2022	13/06/2022
									A070504	Premature Discharge of Battery	13/06/2022	13/06/2022
							F08	Hospitalization or Prolonged Hospitalization	A0705	Battery Problem	13/06/2022	13/06/2022
									A070504	Premature Discharge of Battery	13/06/2022	13/06/2022
							F1905	Device Revision or Replacement	A0705	Battery Problem	13/06/2022	13/06/2022
									A070504	Premature Discharge of Battery	13/06/2022	13/06/2022
					CT449	Generators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem	13/06/2022	13/06/2022
									A070504	Premature Discharge of Battery	13/06/2022	13/06/2022
							F08	Hospitalization or Prolonged Hospitalization	A0705	Battery Problem	13/06/2022	13/06/2022
									A070504	Premature Discharge of Battery	13/06/2022	13/06/2022
							F1905	Device Revision or Replacement	A0705	Battery Problem	13/06/2022	13/06/2022
									A070504	Premature Discharge of Battery	13/06/2022	13/06/2022
					CT981	Single use	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem	13/06/2022	13/06/2022
									A070504	Premature Discharge of Battery	13/06/2022	13/06/2022
							F08	Hospitalization or Prolonged Hospitalization	A0705	Battery Problem	13/06/2022	13/06/2022
									A070504	Premature Discharge of Battery	13/06/2022	13/06/2022
							F1905	Device Revision or Replacement	A0705	Battery Problem	13/06/2022	13/06/2022
									A070504	Premature Discharge of Battery	13/06/2022	13/06/2022
					CT983	Surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem	13/06/2022	13/06/2022
									A070504	Premature Discharge of Battery	13/06/2022	13/06/2022
							F08	Hospitalization or Prolonged Hospitalization	A0705	Battery Problem	13/06/2022	13/06/2022
									A070504	Premature Discharge of Battery	13/06/2022	13/06/2022
							F1905	Device Revision or Replacement	A0705	Battery Problem	13/06/2022	13/06/2022
									A070504	Premature Discharge of Battery	13/06/2022	13/06/2022
6439	2021	All other reportable incidents	35852	Single-chamber implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	21/11/2021	21/11/2021
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	21/11/2021	21/11/2021
							F1903	Device Explantation	A070504	Premature Discharge of Battery	21/11/2021	21/11/2021
6440	2022	Death	47267	Single-chamber implantable pacemaker, rate-responsive	CT141	Electrodes/Leads and associated devices	E0605	Cardiac Tamponade	A24	No Apparent Device Problem	13/04/2022	13/04/2022
							E0619	Pericardial Effusion	A24	No Apparent Device Problem	13/04/2022	13/04/2022
							E2006	Erosion	A24	No Apparent Device Problem	13/04/2022	13/04/2022
							E2114	Perforation	A24	No Apparent Device Problem	13/04/2022	13/04/2022
							E2321	Low Blood Pressure/ Hypotension	A24	No Apparent Device Problem	13/04/2022	13/04/2022
							F02	Death	A24	No Apparent Device Problem	13/04/2022	13/04/2022
					CT145	Cardiovascular devices	E0605	Cardiac Tamponade	A24	No Apparent Device Problem	13/04/2022	13/04/2022
							E0619	Pericardial Effusion	A24	No Apparent Device Problem	13/04/2022	13/04/2022
							E2006	Erosion	A24	No Apparent Device Problem	13/04/2022	13/04/2022
							E2114	Perforation	A24	No Apparent Device Problem	13/04/2022	13/04/2022
							E2321	Low Blood Pressure/ Hypotension	A24	No Apparent Device Problem	13/04/2022	13/04/2022
							F02	Death	A24	No Apparent Device Problem	13/04/2022	13/04/2022
					CT1684	Cardiac pacemakers and associated devices	E0605	Cardiac Tamponade	A24	No Apparent Device Problem	13/04/2022	13/04/2022
							E0619	Pericardial Effusion	A24	No Apparent Device Problem	13/04/2022	13/04/2022
							E2006	Erosion	A24	No Apparent Device Problem	13/04/2022	13/04/2022
							E2114	Perforation	A24	No Apparent Device Problem	13/04/2022	13/04/2022
							E2321	Low Blood Pressure/ Hypotension	A24	No Apparent Device Problem	13/04/2022	13/04/2022
							F02	Death	A24	No Apparent Device Problem	13/04/2022	13/04/2022
					CT1685	Cardiac pacemakers	E0605	Cardiac Tamponade	A24	No Apparent Device Problem	13/04/2022	13/04/2022
							E0619	Pericardial Effusion	A24	No Apparent Device Problem	13/04/2022	13/04/2022
							E2006	Erosion	A24	No Apparent Device Problem	13/04/2022	13/04/2022
							E2114	Perforation	A24	No Apparent Device Problem	13/04/2022	13/04/2022
							E2321	Low Blood Pressure/ Hypotension	A24	No Apparent Device Problem	13/04/2022	13/04/2022
							F02	Death	A24	No Apparent Device Problem	13/04/2022	13/04/2022
					CT233	Surgical	E0605	Cardiac Tamponade	A24	No Apparent Device Problem	13/04/2022	13/04/2022
							E0619	Pericardial Effusion	A24	No Apparent Device Problem	13/04/2022	13/04/2022
							E2006	Erosion	A24	No Apparent Device Problem	13/04/2022	13/04/2022
							E2114	Perforation	A24	No Apparent Device Problem	13/04/2022	13/04/2022
							E2321	Low Blood Pressure/ Hypotension	A24	No Apparent Device Problem	13/04/2022	13/04/2022
							F02	Death	A24	No Apparent Device Problem	13/04/2022	13/04/2022
					CT2406	Implantable	E0605	Cardiac Tamponade	A24	No Apparent Device Problem	13/04/2022	13/04/2022
							E0619	Pericardial Effusion	A24	No Apparent Device Problem	13/04/2022	13/04/2022
							E2006	Erosion	A24	No Apparent Device Problem	13/04/2022	13/04/2022
							E2114	Perforation	A24	No Apparent Device Problem	13/04/2022	13/04/2022
							E2321	Low Blood Pressure/ Hypotension	A24	No Apparent Device Problem	13/04/2022	13/04/2022
							F02	Death	A24	No Apparent Device Problem	13/04/2022	13/04/2022
					CT2407	Active implantable	E0605	Cardiac Tamponade	A24	No Apparent Device Problem	13/04/2022	13/04/2022
							E0619	Pericardial Effusion	A24	No Apparent Device Problem	13/04/2022	13/04/2022
							E2006	Erosion	A24	No Apparent Device Problem	13/04/2022	13/04/2022
							E2114	Perforation	A24	No Apparent Device Problem	13/04/2022	13/04/2022
							E2321	Low Blood Pressure/ Hypotension	A24	No Apparent Device Problem	13/04/2022	13/04/2022
							F02	Death	A24	No Apparent Device Problem	13/04/2022	13/04/2022
					CT257	Implantable cardiac pacemakers	E0605	Cardiac Tamponade	A24	No Apparent Device Problem	13/04/2022	13/04/2022
							E0619	Pericardial Effusion	A24	No Apparent Device Problem	13/04/2022	13/04/2022
							E2006	Erosion	A24	No Apparent Device Problem	13/04/2022	13/04/2022
							E2114	Perforation	A24	No Apparent Device Problem	13/04/2022	13/04/2022
							E2321	Low Blood Pressure/ Hypotension	A24	No Apparent Device Problem	13/04/2022	13/04/2022
							F02	Death	A24	No Apparent Device Problem	13/04/2022	13/04/2022
					CT2727	Active implant devices/ancillaries	E0605	Cardiac Tamponade	A24	No Apparent Device Problem	13/04/2022	13/04/2022
							E0619	Pericardial Effusion	A24	No Apparent Device Problem	13/04/2022	13/04/2022
							E2006	Erosion	A24	No Apparent Device Problem	13/04/2022	13/04/2022
							E2114	Perforation	A24	No Apparent Device Problem	13/04/2022	13/04/2022
							E2321	Low Blood Pressure/ Hypotension	A24	No Apparent Device Problem	13/04/2022	13/04/2022
							F02	Death	A24	No Apparent Device Problem	13/04/2022	13/04/2022

					CT2750	Implantable cardiac electrical stimulators and associated devices	E0605	Cardiac Tamponade	A24	No Apparent Device Problem	13/04/2022	13/04/2022
							E0619	Pericardial Effusion	A24	No Apparent Device Problem	13/04/2022	13/04/2022
							E2006	Erosion	A24	No Apparent Device Problem	13/04/2022	13/04/2022
							E2114	Perforation	A24	No Apparent Device Problem	13/04/2022	13/04/2022
							E2321	Low Blood Pressure/ Hypotension	A24	No Apparent Device Problem	13/04/2022	13/04/2022
							F02	Death	A24	No Apparent Device Problem	13/04/2022	13/04/2022
					CT2866	Implantable cardiac electrical stimulators	E0605	Cardiac Tamponade	A24	No Apparent Device Problem	13/04/2022	13/04/2022
							E0619	Pericardial Effusion	A24	No Apparent Device Problem	13/04/2022	13/04/2022
							E2006	Erosion	A24	No Apparent Device Problem	13/04/2022	13/04/2022
							E2114	Perforation	A24	No Apparent Device Problem	13/04/2022	13/04/2022
							E2321	Low Blood Pressure/ Hypotension	A24	No Apparent Device Problem	13/04/2022	13/04/2022
							F02	Death	A24	No Apparent Device Problem	13/04/2022	13/04/2022
					CT2985	Active	E0605	Cardiac Tamponade	A24	No Apparent Device Problem	13/04/2022	13/04/2022
							E0619	Pericardial Effusion	A24	No Apparent Device Problem	13/04/2022	13/04/2022
							E2006	Erosion	A24	No Apparent Device Problem	13/04/2022	13/04/2022
							E2114	Perforation	A24	No Apparent Device Problem	13/04/2022	13/04/2022
							E2321	Low Blood Pressure/ Hypotension	A24	No Apparent Device Problem	13/04/2022	13/04/2022
							F02	Death	A24	No Apparent Device Problem	13/04/2022	13/04/2022
					CT3001	Implantable metal	E0605	Cardiac Tamponade	A24	No Apparent Device Problem	13/04/2022	13/04/2022
							E0619	Pericardial Effusion	A24	No Apparent Device Problem	13/04/2022	13/04/2022
							E2006	Erosion	A24	No Apparent Device Problem	13/04/2022	13/04/2022
							E2114	Perforation	A24	No Apparent Device Problem	13/04/2022	13/04/2022
							E2321	Low Blood Pressure/ Hypotension	A24	No Apparent Device Problem	13/04/2022	13/04/2022
							F02	Death	A24	No Apparent Device Problem	13/04/2022	13/04/2022
					CT3002	Implantable synthetic polymer	E0605	Cardiac Tamponade	A24	No Apparent Device Problem	13/04/2022	13/04/2022
							E0619	Pericardial Effusion	A24	No Apparent Device Problem	13/04/2022	13/04/2022
							E2006	Erosion	A24	No Apparent Device Problem	13/04/2022	13/04/2022
							E2114	Perforation	A24	No Apparent Device Problem	13/04/2022	13/04/2022
							E2321	Low Blood Pressure/ Hypotension	A24	No Apparent Device Problem	13/04/2022	13/04/2022
							F02	Death	A24	No Apparent Device Problem	13/04/2022	13/04/2022
					CT3006	Implantable inorganic material	E0605	Cardiac Tamponade	A24	No Apparent Device Problem	13/04/2022	13/04/2022
							E0619	Pericardial Effusion	A24	No Apparent Device Problem	13/04/2022	13/04/2022
							E2006	Erosion	A24	No Apparent Device Problem	13/04/2022	13/04/2022
							E2114	Perforation	A24	No Apparent Device Problem	13/04/2022	13/04/2022
							E2321	Low Blood Pressure/ Hypotension	A24	No Apparent Device Problem	13/04/2022	13/04/2022
							F02	Death	A24	No Apparent Device Problem	13/04/2022	13/04/2022
					CT314	Electrically powered	E0605	Cardiac Tamponade	A24	No Apparent Device Problem	13/04/2022	13/04/2022
							E0619	Pericardial Effusion	A24	No Apparent Device Problem	13/04/2022	13/04/2022
							E2006	Erosion	A24	No Apparent Device Problem	13/04/2022	13/04/2022
							E2114	Perforation	A24	No Apparent Device Problem	13/04/2022	13/04/2022
							E2321	Low Blood Pressure/ Hypotension	A24	No Apparent Device Problem	13/04/2022	13/04/2022
							F02	Death	A24	No Apparent Device Problem	13/04/2022	13/04/2022
					CT321	Long-term surgical invasive	E0605	Cardiac Tamponade	A24	No Apparent Device Problem	13/04/2022	13/04/2022
							E0619	Pericardial Effusion	A24	No Apparent Device Problem	13/04/2022	13/04/2022
							E2006	Erosion	A24	No Apparent Device Problem	13/04/2022	13/04/2022
							E2114	Perforation	A24	No Apparent Device Problem	13/04/2022	13/04/2022
							E2321	Low Blood Pressure/ Hypotension	A24	No Apparent Device Problem	13/04/2022	13/04/2022
							F02	Death	A24	No Apparent Device Problem	13/04/2022	13/04/2022
					CT449	Generators and associated devices	E0605	Cardiac Tamponade	A24	No Apparent Device Problem	13/04/2022	13/04/2022
							E0619	Pericardial Effusion	A24	No Apparent Device Problem	13/04/2022	13/04/2022
							E2006	Erosion	A24	No Apparent Device Problem	13/04/2022	13/04/2022
							E2114	Perforation	A24	No Apparent Device Problem	13/04/2022	13/04/2022
							E2321	Low Blood Pressure/ Hypotension	A24	No Apparent Device Problem	13/04/2022	13/04/2022
							F02	Death	A24	No Apparent Device Problem	13/04/2022	13/04/2022
					CT981	Single use	E0605	Cardiac Tamponade	A24	No Apparent Device Problem	13/04/2022	13/04/2022
							E0619	Pericardial Effusion	A24	No Apparent Device Problem	13/04/2022	13/04/2022
							E2006	Erosion	A24	No Apparent Device Problem	13/04/2022	13/04/2022
							E2114	Perforation	A24	No Apparent Device Problem	13/04/2022	13/04/2022
							E2321	Low Blood Pressure/ Hypotension	A24	No Apparent Device Problem	13/04/2022	13/04/2022
							F02	Death	A24	No Apparent Device Problem	13/04/2022	13/04/2022
					CT983	Surgical invasive	E0605	Cardiac Tamponade	A24	No Apparent Device Problem	13/04/2022	13/04/2022
							E0619	Pericardial Effusion	A24	No Apparent Device Problem	13/04/2022	13/04/2022
							E2006	Erosion	A24	No Apparent Device Problem	13/04/2022	13/04/2022
							E2114	Perforation	A24	No Apparent Device Problem	13/04/2022	13/04/2022
							E2321	Low Blood Pressure/ Hypotension	A24	No Apparent Device Problem	13/04/2022	13/04/2022
							F02	Death	A24	No Apparent Device Problem	13/04/2022	13/04/2022
6441	2021	All other reportable incidents	35223	Endocardial/interventricular septal pacing lead	CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A0401	Break		
									A070909	Oversensing		
									A072202	Low Impedance		
							F1901	Additional Surgery	A0401	Break		
									A070909	Oversensing		
									A072202	Low Impedance		
6442	2022	All other reportable incidents	37265	Dual-chamber implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A070909	Oversensing		
							F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A070909	Oversensing		
6443	2022	All other reportable incidents	17882	Physiologic-monitoring defibrillation system	CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070903	Failure to Analyze Signal		
							F26	No Health Consequences or Impact	A070903	Failure to Analyze Signal		
6444	2022	All other reportable incidents	47267	Single-chamber implantable pacemaker, rate-responsive	CT141	Electrodes/Leads and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	30/05/2022	30/05/2022
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	30/05/2022	30/05/2022
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	30/05/2022	30/05/2022
					CT145	Cardiovascular devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	30/05/2022	30/05/2022
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	30/05/2022	30/05/2022
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	30/05/2022	30/05/2022
					CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	30/05/2022	30/05/2022
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	30/05/2022	30/05/2022
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	30/05/2022	30/05/2022
					CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	30/05/2022	30/05/2022
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	30/05/2022	30/05/2022
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	30/05/2022	30/05/2022

					CT233	Surgical	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	30/05/2022	30/05/2022
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	30/05/2022	30/05/2022
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	30/05/2022	30/05/2022
					CT2406	Implantable	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	30/05/2022	30/05/2022
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	30/05/2022	30/05/2022
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	30/05/2022	30/05/2022
					CT2407	Active implantable	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	30/05/2022	30/05/2022
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	30/05/2022	30/05/2022
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	30/05/2022	30/05/2022
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	30/05/2022	30/05/2022
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	30/05/2022	30/05/2022
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	30/05/2022	30/05/2022
					CT2727	Active implant devices/ancillaries	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	30/05/2022	30/05/2022
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	30/05/2022	30/05/2022
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	30/05/2022	30/05/2022
					CT2750	Implantable cardiac electrical stimulators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	30/05/2022	30/05/2022
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	30/05/2022	30/05/2022
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	30/05/2022	30/05/2022
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	30/05/2022	30/05/2022
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	30/05/2022	30/05/2022
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	30/05/2022	30/05/2022
					CT2985	Active	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	30/05/2022	30/05/2022
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	30/05/2022	30/05/2022
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	30/05/2022	30/05/2022
					CT3001	Implantable metal	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	30/05/2022	30/05/2022
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	30/05/2022	30/05/2022
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	30/05/2022	30/05/2022
					CT3002	Implantable synthetic polymer	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	30/05/2022	30/05/2022
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	30/05/2022	30/05/2022
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	30/05/2022	30/05/2022
					CT3006	Implantable inorganic material	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	30/05/2022	30/05/2022
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	30/05/2022	30/05/2022
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	30/05/2022	30/05/2022
					CT314	Electrically powered	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	30/05/2022	30/05/2022
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	30/05/2022	30/05/2022
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	30/05/2022	30/05/2022
					CT321	Long-term surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	30/05/2022	30/05/2022
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	30/05/2022	30/05/2022
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	30/05/2022	30/05/2022
					CT449	Generators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	30/05/2022	30/05/2022
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	30/05/2022	30/05/2022
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	30/05/2022	30/05/2022
					CT981	Single use	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	30/05/2022	30/05/2022
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	30/05/2022	30/05/2022
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	30/05/2022	30/05/2022
					CT983	Surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	30/05/2022	30/05/2022
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	30/05/2022	30/05/2022
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	30/05/2022	30/05/2022
6445	2022	All other reportable incidents	47263	Cardiac resynchronization therapy implantable pacemaker	CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A0908	Incorrect, Inadequate or Imprecise Result or Readings	13/04/2022	13/04/2022
							F08	Hospitalization or Prolonged Hospitalization	A0908	Incorrect, Inadequate or Imprecise Result or Readings	13/04/2022	13/04/2022
							F1905	Device Revision or Replacement	A0908	Incorrect, Inadequate or Imprecise Result or Readings	13/04/2022	13/04/2022
					CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A0908	Incorrect, Inadequate or Imprecise Result or Readings	13/04/2022	13/04/2022
							F08	Hospitalization or Prolonged Hospitalization	A0908	Incorrect, Inadequate or Imprecise Result or Readings	13/04/2022	13/04/2022
							F1905	Device Revision or Replacement	A0908	Incorrect, Inadequate or Imprecise Result or Readings	13/04/2022	13/04/2022
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A0908	Incorrect, Inadequate or Imprecise Result or Readings	13/04/2022	13/04/2022
							F08	Hospitalization or Prolonged Hospitalization	A0908	Incorrect, Inadequate or Imprecise Result or Readings	13/04/2022	13/04/2022
							F1905	Device Revision or Replacement	A0908	Incorrect, Inadequate or Imprecise Result or Readings	13/04/2022	13/04/2022
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A0908	Incorrect, Inadequate or Imprecise Result or Readings	13/04/2022	13/04/2022
							F08	Hospitalization or Prolonged Hospitalization	A0908	Incorrect, Inadequate or Imprecise Result or Readings	13/04/2022	13/04/2022
							F1905	Device Revision or Replacement	A0908	Incorrect, Inadequate or Imprecise Result or Readings	13/04/2022	13/04/2022
6446	2021	All other reportable incidents	17882	Physiologic-monitoring defibrillation system	CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070802	Intermittent Loss of Power		
							F26	No Health Consequences or Impact	A070802	Intermittent Loss of Power		
6447	2022	All other reportable incidents	37265	Dual-chamber implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A072201	High Impedance		
							F26	No Health Consequences or Impact	A072201	High Impedance		
6448	2022	All other reportable incidents	36241	Pacemaker/defibrillator lead	CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A072201	High Impedance		
							F26	No Health Consequences or Impact	A072201	High Impedance		
6449	2022	All other reportable incidents	47267	Single-chamber implantable pacemaker, rate-responsive	CT141	Electrodes/Leads and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem	26/05/2022	26/05/2022
									A070504	Premature Discharge of Battery	26/05/2022	26/05/2022
							F08	Hospitalization or Prolonged Hospitalization	A0705	Battery Problem	26/05/2022	26/05/2022
									A070504	Premature Discharge of Battery	26/05/2022	26/05/2022
							F1905	Device Revision or Replacement	A0705	Battery Problem	26/05/2022	26/05/2022
									A070504	Premature Discharge of Battery	26/05/2022	26/05/2022
					CT145	Cardiovascular devices	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem	26/05/2022	26/05/2022
									A070504	Premature Discharge of Battery	26/05/2022	26/05/2022
							F08	Hospitalization or Prolonged Hospitalization	A0705	Battery Problem	26/05/2022	26/05/2022
									A070504	Premature Discharge of Battery	26/05/2022	26/05/2022
							F1905	Device Revision or Replacement	A0705	Battery Problem	26/05/2022	26/05/2022
									A070504	Premature Discharge of Battery	26/05/2022	26/05/2022
					CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem	26/05/2022	26/05/2022

								A070504	Premature Discharge of Battery	26/05/2022	26/05/2022
						F08	Hospitalization or Prolonged Hospitalization	A0705	Battery Problem	26/05/2022	26/05/2022
								A070504	Premature Discharge of Battery	26/05/2022	26/05/2022
						F1905	Device Revision or Replacement	A0705	Battery Problem	26/05/2022	26/05/2022
								A070504	Premature Discharge of Battery	26/05/2022	26/05/2022
					CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem	26/05/2022	26/05/2022
									A070504	Premature Discharge of Battery	26/05/2022	26/05/2022
							F08	Hospitalization or Prolonged Hospitalization	A0705	Battery Problem	26/05/2022	26/05/2022
									A070504	Premature Discharge of Battery	26/05/2022	26/05/2022
							F1905	Device Revision or Replacement	A0705	Battery Problem	26/05/2022	26/05/2022
									A070504	Premature Discharge of Battery	26/05/2022	26/05/2022
					CT233	Surgical	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem	26/05/2022	26/05/2022
									A070504	Premature Discharge of Battery	26/05/2022	26/05/2022
							F08	Hospitalization or Prolonged Hospitalization	A0705	Battery Problem	26/05/2022	26/05/2022
									A070504	Premature Discharge of Battery	26/05/2022	26/05/2022
							F1905	Device Revision or Replacement	A0705	Battery Problem	26/05/2022	26/05/2022
									A070504	Premature Discharge of Battery	26/05/2022	26/05/2022
					CT2406	Implantable	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem	26/05/2022	26/05/2022
									A070504	Premature Discharge of Battery	26/05/2022	26/05/2022
							F08	Hospitalization or Prolonged Hospitalization	A0705	Battery Problem	26/05/2022	26/05/2022
									A070504	Premature Discharge of Battery	26/05/2022	26/05/2022
							F1905	Device Revision or Replacement	A0705	Battery Problem	26/05/2022	26/05/2022
									A070504	Premature Discharge of Battery	26/05/2022	26/05/2022
					CT2407	Active implantable	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem	26/05/2022	26/05/2022
									A070504	Premature Discharge of Battery	26/05/2022	26/05/2022
							F08	Hospitalization or Prolonged Hospitalization	A0705	Battery Problem	26/05/2022	26/05/2022
									A070504	Premature Discharge of Battery	26/05/2022	26/05/2022
							F1905	Device Revision or Replacement	A0705	Battery Problem	26/05/2022	26/05/2022
									A070504	Premature Discharge of Battery	26/05/2022	26/05/2022
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem	26/05/2022	26/05/2022
									A070504	Premature Discharge of Battery	26/05/2022	26/05/2022
							F08	Hospitalization or Prolonged Hospitalization	A0705	Battery Problem	26/05/2022	26/05/2022
									A070504	Premature Discharge of Battery	26/05/2022	26/05/2022
							F1905	Device Revision or Replacement	A0705	Battery Problem	26/05/2022	26/05/2022
									A070504	Premature Discharge of Battery	26/05/2022	26/05/2022
					CT2727	Active implant devices/ancillaries	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem	26/05/2022	26/05/2022
									A070504	Premature Discharge of Battery	26/05/2022	26/05/2022
							F08	Hospitalization or Prolonged Hospitalization	A0705	Battery Problem	26/05/2022	26/05/2022
									A070504	Premature Discharge of Battery	26/05/2022	26/05/2022
							F1905	Device Revision or Replacement	A0705	Battery Problem	26/05/2022	26/05/2022
									A070504	Premature Discharge of Battery	26/05/2022	26/05/2022
					CT2750	Implantable cardiac electrical stimulators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem	26/05/2022	26/05/2022
									A070504	Premature Discharge of Battery	26/05/2022	26/05/2022
							F08	Hospitalization or Prolonged Hospitalization	A0705	Battery Problem	26/05/2022	26/05/2022
									A070504	Premature Discharge of Battery	26/05/2022	26/05/2022
							F1905	Device Revision or Replacement	A0705	Battery Problem	26/05/2022	26/05/2022
									A070504	Premature Discharge of Battery	26/05/2022	26/05/2022
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem	26/05/2022	26/05/2022
									A070504	Premature Discharge of Battery	26/05/2022	26/05/2022
							F08	Hospitalization or Prolonged Hospitalization	A0705	Battery Problem	26/05/2022	26/05/2022
									A070504	Premature Discharge of Battery	26/05/2022	26/05/2022
							F1905	Device Revision or Replacement	A0705	Battery Problem	26/05/2022	26/05/2022
									A070504	Premature Discharge of Battery	26/05/2022	26/05/2022
					CT2985	Active	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem	26/05/2022	26/05/2022
									A070504	Premature Discharge of Battery	26/05/2022	26/05/2022
							F08	Hospitalization or Prolonged Hospitalization	A0705	Battery Problem	26/05/2022	26/05/2022
									A070504	Premature Discharge of Battery	26/05/2022	26/05/2022
							F1905	Device Revision or Replacement	A0705	Battery Problem	26/05/2022	26/05/2022
									A070504	Premature Discharge of Battery	26/05/2022	26/05/2022
					CT3001	Implantable metal	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem	26/05/2022	26/05/2022
									A070504	Premature Discharge of Battery	26/05/2022	26/05/2022
							F08	Hospitalization or Prolonged Hospitalization	A0705	Battery Problem	26/05/2022	26/05/2022
									A070504	Premature Discharge of Battery	26/05/2022	26/05/2022
							F1905	Device Revision or Replacement	A0705	Battery Problem	26/05/2022	26/05/2022
									A070504	Premature Discharge of Battery	26/05/2022	26/05/2022
					CT3002	Implantable synthetic polymer	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem	26/05/2022	26/05/2022
									A070504	Premature Discharge of Battery	26/05/2022	26/05/2022
							F08	Hospitalization or Prolonged Hospitalization	A0705	Battery Problem	26/05/2022	26/05/2022
									A070504	Premature Discharge of Battery	26/05/2022	26/05/2022
							F1905	Device Revision or Replacement	A0705	Battery Problem	26/05/2022	26/05/2022
									A070504	Premature Discharge of Battery	26/05/2022	26/05/2022
					CT3006	Implantable inorganic material	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem	26/05/2022	26/05/2022
									A070504	Premature Discharge of Battery	26/05/2022	26/05/2022
							F08	Hospitalization or Prolonged Hospitalization	A0705	Battery Problem	26/05/2022	26/05/2022
									A070504	Premature Discharge of Battery	26/05/2022	26/05/2022
							F1905	Device Revision or Replacement	A0705	Battery Problem	26/05/2022	26/05/2022
									A070504	Premature Discharge of Battery	26/05/2022	26/05/2022
					CT314	Electrically powered	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem	26/05/2022	26/05/2022
									A070504	Premature Discharge of Battery	26/05/2022	26/05/2022
							F08	Hospitalization or Prolonged Hospitalization	A0705	Battery Problem	26/05/2022	26/05/2022
									A070504	Premature Discharge of Battery	26/05/2022	26/05/2022
							F1905	Device Revision or Replacement	A0705	Battery Problem	26/05/2022	26/05/2022
									A070504	Premature Discharge of Battery	26/05/2022	26/05/2022
					CT321	Long-term surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem	26/05/2022	26/05/2022
									A070504	Premature Discharge of Battery	26/05/2022	26/05/2022
							F08	Hospitalization or Prolonged Hospitalization	A0705	Battery Problem	26/05/2022	26/05/2022
									A070504	Premature Discharge of Battery	26/05/2022	26/05/2022
							F1905	Device Revision or Replacement	A0705	Battery Problem	26/05/2022	26/05/2022

								A070504	Premature Discharge of Battery	26/05/2022	26/05/2022
				CT449	Generators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem	26/05/2022	26/05/2022
								A070504	Premature Discharge of Battery	26/05/2022	26/05/2022
						F08	Hospitalization or Prolonged Hospitalization	A0705	Battery Problem	26/05/2022	26/05/2022
								A070504	Premature Discharge of Battery	26/05/2022	26/05/2022
						F1905	Device Revision or Replacement	A0705	Battery Problem	26/05/2022	26/05/2022
								A070504	Premature Discharge of Battery	26/05/2022	26/05/2022
				CT981	Single use	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem	26/05/2022	26/05/2022
								A070504	Premature Discharge of Battery	26/05/2022	26/05/2022
						F08	Hospitalization or Prolonged Hospitalization	A0705	Battery Problem	26/05/2022	26/05/2022
								A070504	Premature Discharge of Battery	26/05/2022	26/05/2022
						F1905	Device Revision or Replacement	A0705	Battery Problem	26/05/2022	26/05/2022
								A070504	Premature Discharge of Battery	26/05/2022	26/05/2022
				CT983	Surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem	26/05/2022	26/05/2022
								A070504	Premature Discharge of Battery	26/05/2022	26/05/2022
						F08	Hospitalization or Prolonged Hospitalization	A0705	Battery Problem	26/05/2022	26/05/2022
								A070504	Premature Discharge of Battery	26/05/2022	26/05/2022
						F1905	Device Revision or Replacement	A0705	Battery Problem	26/05/2022	26/05/2022
								A070504	Premature Discharge of Battery	26/05/2022	26/05/2022
6450	2022	All other reportable incidents	47265	Dual-chamber implantable pacemaker, rate-responsive	CT141	Electrodes/Leads and associated devices	E0602	Cardiac Arrest	A0712	Pacing Problem	19/04/2022	19/04/2022
							E2403	No Clinical Signs, Symptoms or Conditions	A0712	Pacing Problem	19/04/2022	19/04/2022
							F1203	Life Threatening Illness or Injury	A0712	Pacing Problem	19/04/2022	19/04/2022
					CT145	Cardiovascular devices	E0602	Cardiac Arrest	A0712	Pacing Problem	19/04/2022	19/04/2022
							E2403	No Clinical Signs, Symptoms or Conditions	A0712	Pacing Problem	19/04/2022	19/04/2022
							F1203	Life Threatening Illness or Injury	A0712	Pacing Problem	19/04/2022	19/04/2022
					CT1684	Cardiac pacemakers and associated devices	E0602	Cardiac Arrest	A0712	Pacing Problem	19/04/2022	19/04/2022
							E2403	No Clinical Signs, Symptoms or Conditions	A0712	Pacing Problem	19/04/2022	19/04/2022
							F1203	Life Threatening Illness or Injury	A0712	Pacing Problem	19/04/2022	19/04/2022
					CT1685	Cardiac pacemakers	E0602	Cardiac Arrest	A0712	Pacing Problem	19/04/2022	19/04/2022
							E2403	No Clinical Signs, Symptoms or Conditions	A0712	Pacing Problem	19/04/2022	19/04/2022
							F1203	Life Threatening Illness or Injury	A0712	Pacing Problem	19/04/2022	19/04/2022
					CT233	Surgical	E0602	Cardiac Arrest	A0712	Pacing Problem	19/04/2022	19/04/2022
							E2403	No Clinical Signs, Symptoms or Conditions	A0712	Pacing Problem	19/04/2022	19/04/2022
							F1203	Life Threatening Illness or Injury	A0712	Pacing Problem	19/04/2022	19/04/2022
					CT2406	Implantable	E0602	Cardiac Arrest	A0712	Pacing Problem	19/04/2022	19/04/2022
							E2403	No Clinical Signs, Symptoms or Conditions	A0712	Pacing Problem	19/04/2022	19/04/2022
							F1203	Life Threatening Illness or Injury	A0712	Pacing Problem	19/04/2022	19/04/2022
					CT2407	Active implantable	E0602	Cardiac Arrest	A0712	Pacing Problem	19/04/2022	19/04/2022
							E2403	No Clinical Signs, Symptoms or Conditions	A0712	Pacing Problem	19/04/2022	19/04/2022
							F1203	Life Threatening Illness or Injury	A0712	Pacing Problem	19/04/2022	19/04/2022
					CT257	Implantable cardiac pacemakers	E0602	Cardiac Arrest	A0712	Pacing Problem	19/04/2022	19/04/2022
							E2403	No Clinical Signs, Symptoms or Conditions	A0712	Pacing Problem	19/04/2022	19/04/2022
							F1203	Life Threatening Illness or Injury	A0712	Pacing Problem	19/04/2022	19/04/2022
					CT2727	Active implant devices/ancillaries	E0602	Cardiac Arrest	A0712	Pacing Problem	19/04/2022	19/04/2022
							E2403	No Clinical Signs, Symptoms or Conditions	A0712	Pacing Problem	19/04/2022	19/04/2022
							F1203	Life Threatening Illness or Injury	A0712	Pacing Problem	19/04/2022	19/04/2022
					CT2750	Implantable cardiac electrical stimulators and associated devices	E0602	Cardiac Arrest	A0712	Pacing Problem	19/04/2022	19/04/2022
							E2403	No Clinical Signs, Symptoms or Conditions	A0712	Pacing Problem	19/04/2022	19/04/2022
							F1203	Life Threatening Illness or Injury	A0712	Pacing Problem	19/04/2022	19/04/2022
					CT2866	Implantable cardiac electrical stimulators	E0602	Cardiac Arrest	A0712	Pacing Problem	19/04/2022	19/04/2022
							E2403	No Clinical Signs, Symptoms or Conditions	A0712	Pacing Problem	19/04/2022	19/04/2022
							F1203	Life Threatening Illness or Injury	A0712	Pacing Problem	19/04/2022	19/04/2022
					CT2985	Active	E0602	Cardiac Arrest	A0712	Pacing Problem	19/04/2022	19/04/2022
							E2403	No Clinical Signs, Symptoms or Conditions	A0712	Pacing Problem	19/04/2022	19/04/2022
							F1203	Life Threatening Illness or Injury	A0712	Pacing Problem	19/04/2022	19/04/2022
					CT3001	Implantable metal	E0602	Cardiac Arrest	A0712	Pacing Problem	19/04/2022	19/04/2022
							E2403	No Clinical Signs, Symptoms or Conditions	A0712	Pacing Problem	19/04/2022	19/04/2022
							F1203	Life Threatening Illness or Injury	A0712	Pacing Problem	19/04/2022	19/04/2022
					CT3002	Implantable synthetic polymer	E0602	Cardiac Arrest	A0712	Pacing Problem	19/04/2022	19/04/2022
							E2403	No Clinical Signs, Symptoms or Conditions	A0712	Pacing Problem	19/04/2022	19/04/2022
							F1203	Life Threatening Illness or Injury	A0712	Pacing Problem	19/04/2022	19/04/2022
					CT3006	Implantable inorganic material	E0602	Cardiac Arrest	A0712	Pacing Problem	19/04/2022	19/04/2022
							E2403	No Clinical Signs, Symptoms or Conditions	A0712	Pacing Problem	19/04/2022	19/04/2022
							F1203	Life Threatening Illness or Injury	A0712	Pacing Problem	19/04/2022	19/04/2022
					CT314	Electrically powered	E0602	Cardiac Arrest	A0712	Pacing Problem	19/04/2022	19/04/2022
							E2403	No Clinical Signs, Symptoms or Conditions	A0712	Pacing Problem	19/04/2022	19/04/2022
							F1203	Life Threatening Illness or Injury	A0712	Pacing Problem	19/04/2022	19/04/2022
					CT321	Long-term surgical invasive	E0602	Cardiac Arrest	A0712	Pacing Problem	19/04/2022	19/04/2022
							E2403	No Clinical Signs, Symptoms or Conditions	A0712	Pacing Problem	19/04/2022	19/04/2022
							F1203	Life Threatening Illness or Injury	A0712	Pacing Problem	19/04/2022	19/04/2022
					CT449	Generators and associated devices	E0602	Cardiac Arrest	A0712	Pacing Problem	19/04/2022	19/04/2022
							E2403	No Clinical Signs, Symptoms or Conditions	A0712	Pacing Problem	19/04/2022	19/04/2022
							F1203	Life Threatening Illness or Injury	A0712	Pacing Problem	19/04/2022	19/04/2022
					CT981	Single use	E0602	Cardiac Arrest	A0712	Pacing Problem	19/04/2022	19/04/2022
							E2403	No Clinical Signs, Symptoms or Conditions	A0712	Pacing Problem	19/04/2022	19/04/2022
							F1203	Life Threatening Illness or Injury	A0712	Pacing Problem	19/04/2022	19/04/2022
					CT983	Surgical invasive	E0602	Cardiac Arrest	A0712	Pacing Problem	19/04/2022	19/04/2022
							E2403	No Clinical Signs, Symptoms or Conditions	A0712	Pacing Problem	19/04/2022	19/04/2022
							F1203	Life Threatening Illness or Injury	A0712	Pacing Problem	19/04/2022	19/04/2022
6451	2021	All other reportable incidents	35852	Single-chamber implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	21/10/2021	21/10/2021
							F19	Surgical Intervention	A070504	Premature Discharge of Battery	21/10/2021	21/10/2021
6452	2022	All other reportable incidents	47270	Cardiac resynchronization therapy implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E0112	Dizziness	A070909	Oversensing		
							E2403	No Clinical Signs, Symptoms or Conditions	A070909	Oversensing		
							F23	Unexpected Medical Intervention	A070909	Oversensing		
6453	2021	All other reportable incidents	35852	Single-chamber implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A0908	Incorrect, Inadequate or Imprecise Result or Readings		
							F13	Misdiagnosis/ Misclassification	A0908	Incorrect, Inadequate or Imprecise Result or Readings		

6454	2022	All other reportable incidents	37265	Dual-chamber implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E0601	Arrhythmia	A070101	Failure to Capture		
									A071206	Pacing Asynchronously		
							E2403	No Clinical Signs, Symptoms or Conditions	A070101	Failure to Capture		
									A071206	Pacing Asynchronously		
							F11	Non-Serious Injury/ Illness/ Impairment	A070101	Failure to Capture		
									A071206	Pacing Asynchronously		
6455	2022	All other reportable incidents	36241	Pacemaker/defibrillator lead	CT1684	Cardiac pacemakers and associated devices	E2008	Foreign Body In Patient	A072201	High Impedance		
							E2403	No Clinical Signs, Symptoms or Conditions	A072201	High Impedance		
							F08	Hospitalization or Prolonged Hospitalization	A072201	High Impedance		
							F1905	Device Revision or Replacement	A072201	High Impedance		
6456	2022	All other reportable incidents	35223	Endocardial/interventricular septal pacing lead	CT1684	Cardiac pacemakers and associated devices	E1906	Unspecified Infection	A24	No Apparent Device Problem	29/05/2015	29/05/2015
							E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	29/05/2015	29/05/2015
							F08	Hospitalization or Prolonged Hospitalization	A24	No Apparent Device Problem	29/05/2015	29/05/2015
							F12	Serious Injury/ Illness/ Impairment	A24	No Apparent Device Problem	29/05/2015	29/05/2015
							F1905	Device Revision or Replacement	A24	No Apparent Device Problem	29/05/2015	29/05/2015
6457	2022	All other reportable incidents	47265	Dual-chamber implantable pacemaker, rate-responsive	CT141	Electrodes/Leads and associated devices	E1906	Unspecified Infection	A24	No Apparent Device Problem	29/05/2015	29/05/2015
							E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	29/05/2015	29/05/2015
							F08	Hospitalization or Prolonged Hospitalization	A24	No Apparent Device Problem	29/05/2015	29/05/2015
							F12	Serious Injury/ Illness/ Impairment	A24	No Apparent Device Problem	29/05/2015	29/05/2015
							F1905	Device Revision or Replacement	A24	No Apparent Device Problem	29/05/2015	29/05/2015
					CT145	Cardiovascular devices	E1906	Unspecified Infection	A24	No Apparent Device Problem	29/05/2015	29/05/2015
							E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	29/05/2015	29/05/2015
							F08	Hospitalization or Prolonged Hospitalization	A24	No Apparent Device Problem	29/05/2015	29/05/2015
							F12	Serious Injury/ Illness/ Impairment	A24	No Apparent Device Problem	29/05/2015	29/05/2015
							F1905	Device Revision or Replacement	A24	No Apparent Device Problem	29/05/2015	29/05/2015
					CT1684	Cardiac pacemakers and associated devices	E1906	Unspecified Infection	A24	No Apparent Device Problem	29/05/2015	29/05/2015
							E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	29/05/2015	29/05/2015
							F08	Hospitalization or Prolonged Hospitalization	A24	No Apparent Device Problem	29/05/2015	29/05/2015
							F12	Serious Injury/ Illness/ Impairment	A24	No Apparent Device Problem	29/05/2015	29/05/2015
							F1905	Device Revision or Replacement	A24	No Apparent Device Problem	29/05/2015	29/05/2015
					CT1685	Cardiac pacemakers	E1906	Unspecified Infection	A24	No Apparent Device Problem	29/05/2015	29/05/2015
							E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	29/05/2015	29/05/2015
							F08	Hospitalization or Prolonged Hospitalization	A24	No Apparent Device Problem	29/05/2015	29/05/2015
							F12	Serious Injury/ Illness/ Impairment	A24	No Apparent Device Problem	29/05/2015	29/05/2015
							F1905	Device Revision or Replacement	A24	No Apparent Device Problem	29/05/2015	29/05/2015
					CT233	Surgical	E1906	Unspecified Infection	A24	No Apparent Device Problem	29/05/2015	29/05/2015
							E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	29/05/2015	29/05/2015
							F08	Hospitalization or Prolonged Hospitalization	A24	No Apparent Device Problem	29/05/2015	29/05/2015
							F12	Serious Injury/ Illness/ Impairment	A24	No Apparent Device Problem	29/05/2015	29/05/2015
							F1905	Device Revision or Replacement	A24	No Apparent Device Problem	29/05/2015	29/05/2015
					CT2406	Implantable	E1906	Unspecified Infection	A24	No Apparent Device Problem	29/05/2015	29/05/2015
							E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	29/05/2015	29/05/2015
							F08	Hospitalization or Prolonged Hospitalization	A24	No Apparent Device Problem	29/05/2015	29/05/2015
							F12	Serious Injury/ Illness/ Impairment	A24	No Apparent Device Problem	29/05/2015	29/05/2015
							F1905	Device Revision or Replacement	A24	No Apparent Device Problem	29/05/2015	29/05/2015
					CT2407	Active implantable	E1906	Unspecified Infection	A24	No Apparent Device Problem	29/05/2015	29/05/2015
							E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	29/05/2015	29/05/2015
							F08	Hospitalization or Prolonged Hospitalization	A24	No Apparent Device Problem	29/05/2015	29/05/2015
							F12	Serious Injury/ Illness/ Impairment	A24	No Apparent Device Problem	29/05/2015	29/05/2015
							F1905	Device Revision or Replacement	A24	No Apparent Device Problem	29/05/2015	29/05/2015
					CT257	Implantable cardiac pacemakers	E1906	Unspecified Infection	A24	No Apparent Device Problem	29/05/2015	29/05/2015
							E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	29/05/2015	29/05/2015
							F08	Hospitalization or Prolonged Hospitalization	A24	No Apparent Device Problem	29/05/2015	29/05/2015
							F12	Serious Injury/ Illness/ Impairment	A24	No Apparent Device Problem	29/05/2015	29/05/2015
							F1905	Device Revision or Replacement	A24	No Apparent Device Problem	29/05/2015	29/05/2015
					CT2727	Active implant devices/ancillaries	E1906	Unspecified Infection	A24	No Apparent Device Problem	29/05/2015	29/05/2015
							E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	29/05/2015	29/05/2015
							F08	Hospitalization or Prolonged Hospitalization	A24	No Apparent Device Problem	29/05/2015	29/05/2015
							F12	Serious Injury/ Illness/ Impairment	A24	No Apparent Device Problem	29/05/2015	29/05/2015
							F1905	Device Revision or Replacement	A24	No Apparent Device Problem	29/05/2015	29/05/2015
					CT2750	Implantable cardiac electrical stimulators and associated devices	E1906	Unspecified Infection	A24	No Apparent Device Problem	29/05/2015	29/05/2015
							E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	29/05/2015	29/05/2015
							F08	Hospitalization or Prolonged Hospitalization	A24	No Apparent Device Problem	29/05/2015	29/05/2015
							F12	Serious Injury/ Illness/ Impairment	A24	No Apparent Device Problem	29/05/2015	29/05/2015
							F1905	Device Revision or Replacement	A24	No Apparent Device Problem	29/05/2015	29/05/2015
					CT2866	Implantable cardiac electrical stimulators	E1906	Unspecified Infection	A24	No Apparent Device Problem	29/05/2015	29/05/2015
							E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	29/05/2015	29/05/2015
							F08	Hospitalization or Prolonged Hospitalization	A24	No Apparent Device Problem	29/05/2015	29/05/2015
							F12	Serious Injury/ Illness/ Impairment	A24	No Apparent Device Problem	29/05/2015	29/05/2015
							F1905	Device Revision or Replacement	A24	No Apparent Device Problem	29/05/2015	29/05/2015
					CT2985	Active	E1906	Unspecified Infection	A24	No Apparent Device Problem	29/05/2015	29/05/2015
							E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	29/05/2015	29/05/2015
							F08	Hospitalization or Prolonged Hospitalization	A24	No Apparent Device Problem	29/05/2015	29/05/2015
							F12	Serious Injury/ Illness/ Impairment	A24	No Apparent Device Problem	29/05/2015	29/05/2015
							F1905	Device Revision or Replacement	A24	No Apparent Device Problem	29/05/2015	29/05/2015
					CT3001	Implantable metal	E1906	Unspecified Infection	A24	No Apparent Device Problem	29/05/2015	29/05/2015
							E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	29/05/2015	29/05/2015
							F08	Hospitalization or Prolonged Hospitalization	A24	No Apparent Device Problem	29/05/2015	29/05/2015
							F12	Serious Injury/ Illness/ Impairment	A24	No Apparent Device Problem	29/05/2015	29/05/2015
							F1905	Device Revision or Replacement	A24	No Apparent Device Problem	29/05/2015	29/05/2015
					CT3002	Implantable synthetic polymer	E1906	Unspecified Infection	A24	No Apparent Device Problem	29/05/2015	29/05/2015
							E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	29/05/2015	29/05/2015
							F08	Hospitalization or Prolonged Hospitalization	A24	No Apparent Device Problem	29/05/2015	29/05/2015
							F12	Serious Injury/ Illness/ Impairment	A24	No Apparent Device Problem	29/05/2015	29/05/2015
							F1905	Device Revision or Replacement	A24	No Apparent Device Problem	29/05/2015	29/05/2015
					CT3006	Implantable inorganic material	E1906	Unspecified Infection	A24	No Apparent Device Problem	29/05/2015	29/05/2015
							E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	29/05/2015	29/05/2015
							F08	Hospitalization or Prolonged Hospitalization	A24	No Apparent Device Problem	29/05/2015	29/05/2015

							F12	Serious Injury/ Illness/ Impairment	A24	No Apparent Device Problem	29/05/2015	29/05/2015
							F1905	Device Revision or Replacement	A24	No Apparent Device Problem	29/05/2015	29/05/2015
					CT314	Electrically powered	E1906	Unspecified Infection	A24	No Apparent Device Problem	29/05/2015	29/05/2015
							E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	29/05/2015	29/05/2015
							F08	Hospitalization or Prolonged Hospitalization	A24	No Apparent Device Problem	29/05/2015	29/05/2015
							F12	Serious Injury/ Illness/ Impairment	A24	No Apparent Device Problem	29/05/2015	29/05/2015
							F1905	Device Revision or Replacement	A24	No Apparent Device Problem	29/05/2015	29/05/2015
					CT321	Long-term surgical invasive	E1906	Unspecified Infection	A24	No Apparent Device Problem	29/05/2015	29/05/2015
							E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	29/05/2015	29/05/2015
							F08	Hospitalization or Prolonged Hospitalization	A24	No Apparent Device Problem	29/05/2015	29/05/2015
							F12	Serious Injury/ Illness/ Impairment	A24	No Apparent Device Problem	29/05/2015	29/05/2015
							F1905	Device Revision or Replacement	A24	No Apparent Device Problem	29/05/2015	29/05/2015
					CT449	Generators and associated devices	E1906	Unspecified Infection	A24	No Apparent Device Problem	29/05/2015	29/05/2015
							E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	29/05/2015	29/05/2015
							F08	Hospitalization or Prolonged Hospitalization	A24	No Apparent Device Problem	29/05/2015	29/05/2015
							F12	Serious Injury/ Illness/ Impairment	A24	No Apparent Device Problem	29/05/2015	29/05/2015
							F1905	Device Revision or Replacement	A24	No Apparent Device Problem	29/05/2015	29/05/2015
					CT981	Single use	E1906	Unspecified Infection	A24	No Apparent Device Problem	29/05/2015	29/05/2015
							E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	29/05/2015	29/05/2015
							F08	Hospitalization or Prolonged Hospitalization	A24	No Apparent Device Problem	29/05/2015	29/05/2015
							F12	Serious Injury/ Illness/ Impairment	A24	No Apparent Device Problem	29/05/2015	29/05/2015
							F1905	Device Revision or Replacement	A24	No Apparent Device Problem	29/05/2015	29/05/2015
					CT983	Surgical invasive	E1906	Unspecified Infection	A24	No Apparent Device Problem	29/05/2015	29/05/2015
							E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	29/05/2015	29/05/2015
							F08	Hospitalization or Prolonged Hospitalization	A24	No Apparent Device Problem	29/05/2015	29/05/2015
							F12	Serious Injury/ Illness/ Impairment	A24	No Apparent Device Problem	29/05/2015	29/05/2015
							F1905	Device Revision or Replacement	A24	No Apparent Device Problem	29/05/2015	29/05/2015
6458	2022	All other reportable incidents	47265	Dual-chamber implantable pacemaker, rate-responsive	CT141	Electrodes/Leads and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT145	Cardiovascular devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT233	Surgical	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT2406	Implantable	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT2407	Active implantable	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT2727	Active implant devices/ancillaries	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT2750	Implantable cardiac electrical stimulators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT2985	Active	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT3001	Implantable metal	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT3002	Implantable synthetic polymer	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT3006	Implantable inorganic material	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT314	Electrically powered	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT321	Long-term surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT449	Generators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT981	Single use	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT983	Surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
6459	2021	All other reportable incidents	47265	Dual-chamber implantable pacemaker, rate-responsive	CT141	Electrodes/Leads and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem	01/08/2021	01/08/2021
									A070504	Premature Discharge of Battery	01/08/2021	01/08/2021
							F08	Hospitalization or Prolonged Hospitalization	A0705	Battery Problem	01/08/2021	01/08/2021
									A070504	Premature Discharge of Battery	01/08/2021	01/08/2021
							F1905	Device Revision or Replacement	A0705	Battery Problem	01/08/2021	01/08/2021
									A070504	Premature Discharge of Battery	01/08/2021	01/08/2021
					CT145	Cardiovascular devices	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem	01/08/2021	01/08/2021
									A070504	Premature Discharge of Battery	01/08/2021	01/08/2021
							F08	Hospitalization or Prolonged Hospitalization	A0705	Battery Problem	01/08/2021	01/08/2021
									A070504	Premature Discharge of Battery	01/08/2021	01/08/2021
							F1905	Device Revision or Replacement	A0705	Battery Problem	01/08/2021	01/08/2021
									A070504	Premature Discharge of Battery	01/08/2021	01/08/2021
					CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem	01/08/2021	01/08/2021
									A070504	Premature Discharge of Battery	01/08/2021	01/08/2021
							F08	Hospitalization or Prolonged Hospitalization	A0705	Battery Problem	01/08/2021	01/08/2021
									A070504	Premature Discharge of Battery	01/08/2021	01/08/2021
							F1905	Device Revision or Replacement	A0705	Battery Problem	01/08/2021	01/08/2021
									A070504	Premature Discharge of Battery	01/08/2021	01/08/2021
					CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem	01/08/2021	01/08/2021
									A070504	Premature Discharge of Battery	01/08/2021	01/08/2021
							F08	Hospitalization or Prolonged Hospitalization	A0705	Battery Problem	01/08/2021	01/08/2021

								A070504	Premature Discharge of Battery	01/08/2021	01/08/2021
						F1905	Device Revision or Replacement	A0705	Battery Problem	01/08/2021	01/08/2021
								A070504	Premature Discharge of Battery	01/08/2021	01/08/2021
					CT233	Surgical	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem	01/08/2021	01/08/2021
									A070504	Premature Discharge of Battery	01/08/2021	01/08/2021
							F08	Hospitalization or Prolonged Hospitalization	A0705	Battery Problem	01/08/2021	01/08/2021
									A070504	Premature Discharge of Battery	01/08/2021	01/08/2021
							F1905	Device Revision or Replacement	A0705	Battery Problem	01/08/2021	01/08/2021
									A070504	Premature Discharge of Battery	01/08/2021	01/08/2021
					CT2406	Implantable	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem	01/08/2021	01/08/2021
									A070504	Premature Discharge of Battery	01/08/2021	01/08/2021
							F08	Hospitalization or Prolonged Hospitalization	A0705	Battery Problem	01/08/2021	01/08/2021
									A070504	Premature Discharge of Battery	01/08/2021	01/08/2021
							F1905	Device Revision or Replacement	A0705	Battery Problem	01/08/2021	01/08/2021
									A070504	Premature Discharge of Battery	01/08/2021	01/08/2021
					CT2407	Active implantable	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem	01/08/2021	01/08/2021
									A070504	Premature Discharge of Battery	01/08/2021	01/08/2021
							F08	Hospitalization or Prolonged Hospitalization	A0705	Battery Problem	01/08/2021	01/08/2021
									A070504	Premature Discharge of Battery	01/08/2021	01/08/2021
							F1905	Device Revision or Replacement	A0705	Battery Problem	01/08/2021	01/08/2021
									A070504	Premature Discharge of Battery	01/08/2021	01/08/2021
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem	01/08/2021	01/08/2021
									A070504	Premature Discharge of Battery	01/08/2021	01/08/2021
							F08	Hospitalization or Prolonged Hospitalization	A0705	Battery Problem	01/08/2021	01/08/2021
									A070504	Premature Discharge of Battery	01/08/2021	01/08/2021
							F1905	Device Revision or Replacement	A0705	Battery Problem	01/08/2021	01/08/2021
									A070504	Premature Discharge of Battery	01/08/2021	01/08/2021
					CT2727	Active implant devices/ancillaries	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem	01/08/2021	01/08/2021
									A070504	Premature Discharge of Battery	01/08/2021	01/08/2021
							F08	Hospitalization or Prolonged Hospitalization	A0705	Battery Problem	01/08/2021	01/08/2021
									A070504	Premature Discharge of Battery	01/08/2021	01/08/2021
							F1905	Device Revision or Replacement	A0705	Battery Problem	01/08/2021	01/08/2021
									A070504	Premature Discharge of Battery	01/08/2021	01/08/2021
					CT2750	Implantable cardiac electrical stimulators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem	01/08/2021	01/08/2021
									A070504	Premature Discharge of Battery	01/08/2021	01/08/2021
							F08	Hospitalization or Prolonged Hospitalization	A0705	Battery Problem	01/08/2021	01/08/2021
									A070504	Premature Discharge of Battery	01/08/2021	01/08/2021
							F1905	Device Revision or Replacement	A0705	Battery Problem	01/08/2021	01/08/2021
									A070504	Premature Discharge of Battery	01/08/2021	01/08/2021
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem	01/08/2021	01/08/2021
									A070504	Premature Discharge of Battery	01/08/2021	01/08/2021
							F08	Hospitalization or Prolonged Hospitalization	A0705	Battery Problem	01/08/2021	01/08/2021
									A070504	Premature Discharge of Battery	01/08/2021	01/08/2021
							F1905	Device Revision or Replacement	A0705	Battery Problem	01/08/2021	01/08/2021
									A070504	Premature Discharge of Battery	01/08/2021	01/08/2021
					CT2985	Active	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem	01/08/2021	01/08/2021
									A070504	Premature Discharge of Battery	01/08/2021	01/08/2021
							F08	Hospitalization or Prolonged Hospitalization	A0705	Battery Problem	01/08/2021	01/08/2021
									A070504	Premature Discharge of Battery	01/08/2021	01/08/2021
							F1905	Device Revision or Replacement	A0705	Battery Problem	01/08/2021	01/08/2021
									A070504	Premature Discharge of Battery	01/08/2021	01/08/2021
					CT3001	Implantable metal	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem	01/08/2021	01/08/2021
									A070504	Premature Discharge of Battery	01/08/2021	01/08/2021
							F08	Hospitalization or Prolonged Hospitalization	A0705	Battery Problem	01/08/2021	01/08/2021
									A070504	Premature Discharge of Battery	01/08/2021	01/08/2021
							F1905	Device Revision or Replacement	A0705	Battery Problem	01/08/2021	01/08/2021
									A070504	Premature Discharge of Battery	01/08/2021	01/08/2021
					CT3002	Implantable synthetic polymer	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem	01/08/2021	01/08/2021
									A070504	Premature Discharge of Battery	01/08/2021	01/08/2021
							F08	Hospitalization or Prolonged Hospitalization	A0705	Battery Problem	01/08/2021	01/08/2021
									A070504	Premature Discharge of Battery	01/08/2021	01/08/2021
							F1905	Device Revision or Replacement	A0705	Battery Problem	01/08/2021	01/08/2021
									A070504	Premature Discharge of Battery	01/08/2021	01/08/2021
					CT3006	Implantable inorganic material	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem	01/08/2021	01/08/2021
									A070504	Premature Discharge of Battery	01/08/2021	01/08/2021
							F08	Hospitalization or Prolonged Hospitalization	A0705	Battery Problem	01/08/2021	01/08/2021
									A070504	Premature Discharge of Battery	01/08/2021	01/08/2021
							F1905	Device Revision or Replacement	A0705	Battery Problem	01/08/2021	01/08/2021
									A070504	Premature Discharge of Battery	01/08/2021	01/08/2021
					CT314	Electrically powered	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem	01/08/2021	01/08/2021
									A070504	Premature Discharge of Battery	01/08/2021	01/08/2021
							F08	Hospitalization or Prolonged Hospitalization	A0705	Battery Problem	01/08/2021	01/08/2021
									A070504	Premature Discharge of Battery	01/08/2021	01/08/2021
							F1905	Device Revision or Replacement	A0705	Battery Problem	01/08/2021	01/08/2021
									A070504	Premature Discharge of Battery	01/08/2021	01/08/2021
					CT321	Long-term surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem	01/08/2021	01/08/2021
									A070504	Premature Discharge of Battery	01/08/2021	01/08/2021
							F08	Hospitalization or Prolonged Hospitalization	A0705	Battery Problem	01/08/2021	01/08/2021
									A070504	Premature Discharge of Battery	01/08/2021	01/08/2021
							F1905	Device Revision or Replacement	A0705	Battery Problem	01/08/2021	01/08/2021
									A070504	Premature Discharge of Battery	01/08/2021	01/08/2021
					CT449	Generators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem	01/08/2021	01/08/2021
									A070504	Premature Discharge of Battery	01/08/2021	01/08/2021
							F08	Hospitalization or Prolonged Hospitalization	A0705	Battery Problem	01/08/2021	01/08/2021
									A070504	Premature Discharge of Battery	01/08/2021	01/08/2021
							F1905	Device Revision or Replacement	A0705	Battery Problem	01/08/2021	01/08/2021
									A070504	Premature Discharge of Battery	01/08/2021	01/08/2021
					CT981	Single use	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem	01/08/2021	01/08/2021

									A070504	Premature Discharge of Battery	01/08/2021	01/08/2021
							F08	Hospitalization or Prolonged Hospitalization	A0705	Battery Problem	01/08/2021	01/08/2021
									A070504	Premature Discharge of Battery	01/08/2021	01/08/2021
							F1905	Device Revision or Replacement	A0705	Battery Problem	01/08/2021	01/08/2021
									A070504	Premature Discharge of Battery	01/08/2021	01/08/2021
					CT983	Surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem	01/08/2021	01/08/2021
									A070504	Premature Discharge of Battery	01/08/2021	01/08/2021
							F08	Hospitalization or Prolonged Hospitalization	A0705	Battery Problem	01/08/2021	01/08/2021
									A070504	Premature Discharge of Battery	01/08/2021	01/08/2021
							F1905	Device Revision or Replacement	A0705	Battery Problem	01/08/2021	01/08/2021
									A070504	Premature Discharge of Battery	01/08/2021	01/08/2021
6460	2021	All other reportable incidents	47265	Dual-chamber implantable pacemaker, rate-responsive	CT141	Electrodes/Leads and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
					CT145	Cardiovascular devices	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
					CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
					CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
					CT233	Surgical	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
					CT2406	Implantable	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
					CT2407	Active implantable	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
					CT2727	Active implant devices/ancillaries	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
					CT2750	Implantable cardiac electrical stimulators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
					CT2985	Active	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
					CT3001	Implantable metal	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
					CT3002	Implantable synthetic polymer	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
					CT3006	Implantable inorganic material	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
					CT314	Electrically powered	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
					CT321	Long-term surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
					CT449	Generators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
					CT981	Single use	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
					CT983	Surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem		

									A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
6461	2022	All other reportable incidents	36241	Pacemaker/defibrillator lead	CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A072201	High Impedance	16/03/2022	16/03/2022
							F1908	Prolonged Surgery	A072201	High Impedance	16/03/2022	16/03/2022
6462	2021	All other reportable incidents	35852	Single-chamber implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem	13/10/2021	13/10/2021
									A070504	Premature Discharge of Battery	13/10/2021	13/10/2021
							F08	Hospitalization or Prolonged Hospitalization	A0705	Battery Problem	13/10/2021	13/10/2021
									A070504	Premature Discharge of Battery	13/10/2021	13/10/2021
							F1905	Device Revision or Replacement	A0705	Battery Problem	13/10/2021	13/10/2021
									A070504	Premature Discharge of Battery	13/10/2021	13/10/2021
6463	2021	All other reportable incidents	37265	Dual-chamber implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	14/09/2021	14/09/2021
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	14/09/2021	14/09/2021
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	14/09/2021	14/09/2021
6464	2021	All other reportable incidents	47265	Dual-chamber implantable pacemaker, rate-responsive	CT141	Electrodes/Leads and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	05/08/2021	05/08/2021
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	05/08/2021	05/08/2021
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	05/08/2021	05/08/2021
					CT145	Cardiovascular devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	05/08/2021	05/08/2021
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	05/08/2021	05/08/2021
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	05/08/2021	05/08/2021
					CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	05/08/2021	05/08/2021
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	05/08/2021	05/08/2021
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	05/08/2021	05/08/2021
					CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	05/08/2021	05/08/2021
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	05/08/2021	05/08/2021
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	05/08/2021	05/08/2021
					CT233	Surgical	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	05/08/2021	05/08/2021
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	05/08/2021	05/08/2021
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	05/08/2021	05/08/2021
					CT2406	Implantable	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	05/08/2021	05/08/2021
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	05/08/2021	05/08/2021
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	05/08/2021	05/08/2021
					CT2407	Active implantable	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	05/08/2021	05/08/2021
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	05/08/2021	05/08/2021
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	05/08/2021	05/08/2021
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	05/08/2021	05/08/2021
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	05/08/2021	05/08/2021
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	05/08/2021	05/08/2021
					CT2727	Active implant devices/ancillaries	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	05/08/2021	05/08/2021
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	05/08/2021	05/08/2021
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	05/08/2021	05/08/2021
					CT2750	Implantable cardiac electrical stimulators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	05/08/2021	05/08/2021
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	05/08/2021	05/08/2021
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	05/08/2021	05/08/2021
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	05/08/2021	05/08/2021
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	05/08/2021	05/08/2021
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	05/08/2021	05/08/2021
					CT2985	Active	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	05/08/2021	05/08/2021
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	05/08/2021	05/08/2021
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	05/08/2021	05/08/2021
					CT3001	Implantable metal	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	05/08/2021	05/08/2021
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	05/08/2021	05/08/2021
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	05/08/2021	05/08/2021
					CT3002	Implantable synthetic polymer	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	05/08/2021	05/08/2021
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	05/08/2021	05/08/2021
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	05/08/2021	05/08/2021
					CT3006	Implantable inorganic material	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	05/08/2021	05/08/2021
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	05/08/2021	05/08/2021
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	05/08/2021	05/08/2021
					CT314	Electrically powered	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	05/08/2021	05/08/2021
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	05/08/2021	05/08/2021
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	05/08/2021	05/08/2021
					CT321	Long-term surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	05/08/2021	05/08/2021
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	05/08/2021	05/08/2021
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	05/08/2021	05/08/2021
					CT449	Generators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	05/08/2021	05/08/2021
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	05/08/2021	05/08/2021
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	05/08/2021	05/08/2021
					CT981	Single use	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	05/08/2021	05/08/2021
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	05/08/2021	05/08/2021
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	05/08/2021	05/08/2021
					CT983	Surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	05/08/2021	05/08/2021
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	05/08/2021	05/08/2021
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	05/08/2021	05/08/2021
6465	2022	All other reportable incidents	47270	Cardiac resynchronization therapy implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A070101	Failure to Capture		
									A070909	Oversensing		
							F26	No Health Consequences or Impact	A070101	Failure to Capture		
									A070909	Oversensing		
6466	2021	All other reportable incidents	35223	Endocardial/interventricular septal pacing lead	CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A0401	Break		
									A070901	Decreased Sensitivity		
									A070909	Oversensing		
									A070910	Undersensing		
									A0722	Impedance Problem		
									A090801	Signal Artifact/Noise		
							F26	No Health Consequences or Impact	A0401	Break		
									A070901	Decreased Sensitivity		
									A070909	Oversensing		
									A070910	Undersensing		

									A0722	Impedance Problem		
									A090801	Signal Artifact/Noise		
6467	2021	All other reportable incidents	35223	Endocardial/interventricular septal pacing lead	CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A0401	Break		
									A0701	Capturing Problem		
									A070901	Decreased Sensitivity		
									A070909	Oversensing		
									A0722	Impedance Problem		
									A090801	Signal Artifact/Noise		
							F26	No Health Consequences or Impact	A0401	Break		
									A0701	Capturing Problem		
									A070901	Decreased Sensitivity		
									A070909	Oversensing		
									A0722	Impedance Problem		
									A090801	Signal Artifact/Noise		
6468	2022	All other reportable incidents	47265	Dual-chamber implantable pacemaker, rate-responsive	CT141	Electrodes/Leads and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070909	Oversensing		
									A072001	Electromagnetic Interference		
									A090801	Signal Artifact/Noise		
							F26	No Health Consequences or Impact	A070909	Oversensing		
									A072001	Electromagnetic Interference		
									A090801	Signal Artifact/Noise		
					CT145	Cardiovascular devices	E2403	No Clinical Signs, Symptoms or Conditions	A070909	Oversensing		
									A072001	Electromagnetic Interference		
									A090801	Signal Artifact/Noise		
							F26	No Health Consequences or Impact	A070909	Oversensing		
									A072001	Electromagnetic Interference		
									A090801	Signal Artifact/Noise		
					CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070909	Oversensing		
									A072001	Electromagnetic Interference		
									A090801	Signal Artifact/Noise		
							F26	No Health Consequences or Impact	A070909	Oversensing		
									A072001	Electromagnetic Interference		
									A090801	Signal Artifact/Noise		
					CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A070909	Oversensing		
									A072001	Electromagnetic Interference		
									A090801	Signal Artifact/Noise		
							F26	No Health Consequences or Impact	A070909	Oversensing		
									A072001	Electromagnetic Interference		
									A090801	Signal Artifact/Noise		
					CT233	Surgical	E2403	No Clinical Signs, Symptoms or Conditions	A070909	Oversensing		
									A072001	Electromagnetic Interference		
									A090801	Signal Artifact/Noise		
							F26	No Health Consequences or Impact	A070909	Oversensing		
									A072001	Electromagnetic Interference		
									A090801	Signal Artifact/Noise		
					CT2406	Implantable	E2403	No Clinical Signs, Symptoms or Conditions	A070909	Oversensing		
									A072001	Electromagnetic Interference		
									A090801	Signal Artifact/Noise		
							F26	No Health Consequences or Impact	A070909	Oversensing		
									A072001	Electromagnetic Interference		
									A090801	Signal Artifact/Noise		
					CT2407	Active implantable	E2403	No Clinical Signs, Symptoms or Conditions	A070909	Oversensing		
									A072001	Electromagnetic Interference		
									A090801	Signal Artifact/Noise		
							F26	No Health Consequences or Impact	A070909	Oversensing		
									A072001	Electromagnetic Interference		
									A090801	Signal Artifact/Noise		
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A070909	Oversensing		
									A072001	Electromagnetic Interference		
									A090801	Signal Artifact/Noise		
							F26	No Health Consequences or Impact	A070909	Oversensing		
									A072001	Electromagnetic Interference		
									A090801	Signal Artifact/Noise		
					CT2727	Active implant devices/ancillaries	E2403	No Clinical Signs, Symptoms or Conditions	A070909	Oversensing		
									A072001	Electromagnetic Interference		
									A090801	Signal Artifact/Noise		
							F26	No Health Consequences or Impact	A070909	Oversensing		
									A072001	Electromagnetic Interference		
									A090801	Signal Artifact/Noise		
					CT2750	Implantable cardiac electrical stimulators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070909	Oversensing		
									A072001	Electromagnetic Interference		
									A090801	Signal Artifact/Noise		
							F26	No Health Consequences or Impact	A070909	Oversensing		
									A072001	Electromagnetic Interference		
									A090801	Signal Artifact/Noise		
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A070909	Oversensing		
									A072001	Electromagnetic Interference		
									A090801	Signal Artifact/Noise		
							F26	No Health Consequences or Impact	A070909	Oversensing		
									A072001	Electromagnetic Interference		
									A090801	Signal Artifact/Noise		
					CT2985	Active	E2403	No Clinical Signs, Symptoms or Conditions	A070909	Oversensing		
									A072001	Electromagnetic Interference		
									A090801	Signal Artifact/Noise		
							F26	No Health Consequences or Impact	A070909	Oversensing		
									A072001	Electromagnetic Interference		
									A090801	Signal Artifact/Noise		
					CT3001	Implantable metal	E2403	No Clinical Signs, Symptoms or Conditions	A070909	Oversensing		
									A072001	Electromagnetic Interference		

								A090801	Signal Artifact/Noise		
						F26	No Health Consequences or Impact	A070909	Oversensing		
								A072001	Electromagnetic Interference		
								A090801	Signal Artifact/Noise		
					CT3002	Implantable synthetic polymer	E2403	No Clinical Signs, Symptoms or Conditions	A070909	Oversensing		
									A072001	Electromagnetic Interference		
									A090801	Signal Artifact/Noise		
							F26	No Health Consequences or Impact	A070909	Oversensing		
									A072001	Electromagnetic Interference		
									A090801	Signal Artifact/Noise		
					CT3006	Implantable inorganic material	E2403	No Clinical Signs, Symptoms or Conditions	A070909	Oversensing		
									A072001	Electromagnetic Interference		
									A090801	Signal Artifact/Noise		
							F26	No Health Consequences or Impact	A070909	Oversensing		
									A072001	Electromagnetic Interference		
									A090801	Signal Artifact/Noise		
					CT314	Electrically powered	E2403	No Clinical Signs, Symptoms or Conditions	A070909	Oversensing		
									A072001	Electromagnetic Interference		
									A090801	Signal Artifact/Noise		
							F26	No Health Consequences or Impact	A070909	Oversensing		
									A072001	Electromagnetic Interference		
									A090801	Signal Artifact/Noise		
					CT321	Long-term surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A070909	Oversensing		
									A072001	Electromagnetic Interference		
									A090801	Signal Artifact/Noise		
							F26	No Health Consequences or Impact	A070909	Oversensing		
									A072001	Electromagnetic Interference		
									A090801	Signal Artifact/Noise		
					CT449	Generators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070909	Oversensing		
									A072001	Electromagnetic Interference		
									A090801	Signal Artifact/Noise		
							F26	No Health Consequences or Impact	A070909	Oversensing		
									A072001	Electromagnetic Interference		
									A090801	Signal Artifact/Noise		
					CT981	Single use	E2403	No Clinical Signs, Symptoms or Conditions	A070909	Oversensing		
									A072001	Electromagnetic Interference		
									A090801	Signal Artifact/Noise		
							F26	No Health Consequences or Impact	A070909	Oversensing		
									A072001	Electromagnetic Interference		
									A090801	Signal Artifact/Noise		
					CT983	Surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A070909	Oversensing		
									A072001	Electromagnetic Interference		
									A090801	Signal Artifact/Noise		
							F26	No Health Consequences or Impact	A070909	Oversensing		
									A072001	Electromagnetic Interference		
									A090801	Signal Artifact/Noise		
6469	2022	All other reportable incidents	47265	Dual-chamber implantable pacemaker, rate-responsive	CT141	Electrodes/Leads and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070909	Oversensing		
									A072001	Electromagnetic Interference		
									A090801	Signal Artifact/Noise		
							F26	No Health Consequences or Impact	A070909	Oversensing		
									A072001	Electromagnetic Interference		
									A090801	Signal Artifact/Noise		
					CT145	Cardiovascular devices	E2403	No Clinical Signs, Symptoms or Conditions	A070909	Oversensing		
									A072001	Electromagnetic Interference		
									A090801	Signal Artifact/Noise		
							F26	No Health Consequences or Impact	A070909	Oversensing		
									A072001	Electromagnetic Interference		
									A090801	Signal Artifact/Noise		
					CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070909	Oversensing		
									A072001	Electromagnetic Interference		
									A090801	Signal Artifact/Noise		
							F26	No Health Consequences or Impact	A070909	Oversensing		
									A072001	Electromagnetic Interference		
									A090801	Signal Artifact/Noise		
					CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A070909	Oversensing		
									A072001	Electromagnetic Interference		
									A090801	Signal Artifact/Noise		
							F26	No Health Consequences or Impact	A070909	Oversensing		
									A072001	Electromagnetic Interference		
									A090801	Signal Artifact/Noise		
					CT233	Surgical	E2403	No Clinical Signs, Symptoms or Conditions	A070909	Oversensing		
									A072001	Electromagnetic Interference		
									A090801	Signal Artifact/Noise		
							F26	No Health Consequences or Impact	A070909	Oversensing		
									A072001	Electromagnetic Interference		
									A090801	Signal Artifact/Noise		
					CT2406	Implantable	E2403	No Clinical Signs, Symptoms or Conditions	A070909	Oversensing		
									A072001	Electromagnetic Interference		
									A090801	Signal Artifact/Noise		
							F26	No Health Consequences or Impact	A070909	Oversensing		
									A072001	Electromagnetic Interference		
									A090801	Signal Artifact/Noise		
					CT2407	Active implantable	E2403	No Clinical Signs, Symptoms or Conditions	A070909	Oversensing		
									A072001	Electromagnetic Interference		
									A090801	Signal Artifact/Noise		
							F26	No Health Consequences or Impact	A070909	Oversensing		
									A072001	Electromagnetic Interference		
									A090801	Signal Artifact/Noise		

					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A070909	Oversensing		
									A072001	Electromagnetic Interference		
									A090801	Signal Artifact/Noise		
							F26	No Health Consequences or Impact	A070909	Oversensing		
									A072001	Electromagnetic Interference		
									A090801	Signal Artifact/Noise		
					CT2727	Active implant devices/ancillaries	E2403	No Clinical Signs, Symptoms or Conditions	A070909	Oversensing		
									A072001	Electromagnetic Interference		
									A090801	Signal Artifact/Noise		
							F26	No Health Consequences or Impact	A070909	Oversensing		
									A072001	Electromagnetic Interference		
									A090801	Signal Artifact/Noise		
					CT2750	Implantable cardiac electrical stimulators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070909	Oversensing		
									A072001	Electromagnetic Interference		
									A090801	Signal Artifact/Noise		
							F26	No Health Consequences or Impact	A070909	Oversensing		
									A072001	Electromagnetic Interference		
									A090801	Signal Artifact/Noise		
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A070909	Oversensing		
									A072001	Electromagnetic Interference		
									A090801	Signal Artifact/Noise		
							F26	No Health Consequences or Impact	A070909	Oversensing		
									A072001	Electromagnetic Interference		
									A090801	Signal Artifact/Noise		
					CT2985	Active	E2403	No Clinical Signs, Symptoms or Conditions	A070909	Oversensing		
									A072001	Electromagnetic Interference		
									A090801	Signal Artifact/Noise		
							F26	No Health Consequences or Impact	A070909	Oversensing		
									A072001	Electromagnetic Interference		
									A090801	Signal Artifact/Noise		
					CT3001	Implantable metal	E2403	No Clinical Signs, Symptoms or Conditions	A070909	Oversensing		
									A072001	Electromagnetic Interference		
									A090801	Signal Artifact/Noise		
							F26	No Health Consequences or Impact	A070909	Oversensing		
									A072001	Electromagnetic Interference		
									A090801	Signal Artifact/Noise		
					CT3002	Implantable synthetic polymer	E2403	No Clinical Signs, Symptoms or Conditions	A070909	Oversensing		
									A072001	Electromagnetic Interference		
									A090801	Signal Artifact/Noise		
							F26	No Health Consequences or Impact	A070909	Oversensing		
									A072001	Electromagnetic Interference		
									A090801	Signal Artifact/Noise		
					CT3006	Implantable inorganic material	E2403	No Clinical Signs, Symptoms or Conditions	A070909	Oversensing		
									A072001	Electromagnetic Interference		
									A090801	Signal Artifact/Noise		
							F26	No Health Consequences or Impact	A070909	Oversensing		
									A072001	Electromagnetic Interference		
									A090801	Signal Artifact/Noise		
					CT314	Electrically powered	E2403	No Clinical Signs, Symptoms or Conditions	A070909	Oversensing		
									A072001	Electromagnetic Interference		
									A090801	Signal Artifact/Noise		
							F26	No Health Consequences or Impact	A070909	Oversensing		
									A072001	Electromagnetic Interference		
									A090801	Signal Artifact/Noise		
					CT321	Long-term surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A070909	Oversensing		
									A072001	Electromagnetic Interference		
									A090801	Signal Artifact/Noise		
							F26	No Health Consequences or Impact	A070909	Oversensing		
									A072001	Electromagnetic Interference		
									A090801	Signal Artifact/Noise		
					CT449	Generators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070909	Oversensing		
									A072001	Electromagnetic Interference		
									A090801	Signal Artifact/Noise		
							F26	No Health Consequences or Impact	A070909	Oversensing		
									A072001	Electromagnetic Interference		
									A090801	Signal Artifact/Noise		
					CT981	Single use	E2403	No Clinical Signs, Symptoms or Conditions	A070909	Oversensing		
									A072001	Electromagnetic Interference		
									A090801	Signal Artifact/Noise		
							F26	No Health Consequences or Impact	A070909	Oversensing		
									A072001	Electromagnetic Interference		
									A090801	Signal Artifact/Noise		
					CT983	Surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A070909	Oversensing		
									A072001	Electromagnetic Interference		
									A090801	Signal Artifact/Noise		
							F26	No Health Consequences or Impact	A070909	Oversensing		
									A072001	Electromagnetic Interference		
									A090801	Signal Artifact/Noise		
6470	2022	All other reportable incidents	35223	Endocardial/interventricular septal pacing lead	CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A090801	Signal Artifact/Noise		
							F26	No Health Consequences or Impact	A090801	Signal Artifact/Noise		
6471	2022	All other reportable incidents	47265	Dual-chamber implantable pacemaker, rate-responsive	CT141	Electrodes/Leads and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem	22/06/2022	22/06/2022
									A070504	Premature Discharge of Battery	22/06/2022	22/06/2022
							F08	Hospitalization or Prolonged Hospitalization	A0705	Battery Problem	22/06/2022	22/06/2022
									A070504	Premature Discharge of Battery	22/06/2022	22/06/2022
							F1905	Device Revision or Replacement	A0705	Battery Problem	22/06/2022	22/06/2022
									A070504	Premature Discharge of Battery	22/06/2022	22/06/2022
					CT145	Cardiovascular devices	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem	22/06/2022	22/06/2022
									A070504	Premature Discharge of Battery	22/06/2022	22/06/2022

		F08	Hospitalization or Prolonged Hospitalization	A0705	Battery Problem	22/06/2022	22/06/2022
				A070504	Premature Discharge of Battery	22/06/2022	22/06/2022
		F1905	Device Revision or Replacement	A0705	Battery Problem	22/06/2022	22/06/2022
				A070504	Premature Discharge of Battery	22/06/2022	22/06/2022
CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem	22/06/2022	22/06/2022
				A070504	Premature Discharge of Battery	22/06/2022	22/06/2022
		F08	Hospitalization or Prolonged Hospitalization	A0705	Battery Problem	22/06/2022	22/06/2022
				A070504	Premature Discharge of Battery	22/06/2022	22/06/2022
		F1905	Device Revision or Replacement	A0705	Battery Problem	22/06/2022	22/06/2022
				A070504	Premature Discharge of Battery	22/06/2022	22/06/2022
CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem	22/06/2022	22/06/2022
				A070504	Premature Discharge of Battery	22/06/2022	22/06/2022
		F08	Hospitalization or Prolonged Hospitalization	A0705	Battery Problem	22/06/2022	22/06/2022
				A070504	Premature Discharge of Battery	22/06/2022	22/06/2022
		F1905	Device Revision or Replacement	A0705	Battery Problem	22/06/2022	22/06/2022
				A070504	Premature Discharge of Battery	22/06/2022	22/06/2022
CT233	Surgical	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem	22/06/2022	22/06/2022
				A070504	Premature Discharge of Battery	22/06/2022	22/06/2022
		F08	Hospitalization or Prolonged Hospitalization	A0705	Battery Problem	22/06/2022	22/06/2022
				A070504	Premature Discharge of Battery	22/06/2022	22/06/2022
		F1905	Device Revision or Replacement	A0705	Battery Problem	22/06/2022	22/06/2022
				A070504	Premature Discharge of Battery	22/06/2022	22/06/2022
CT2406	Implantable	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem	22/06/2022	22/06/2022
				A070504	Premature Discharge of Battery	22/06/2022	22/06/2022
		F08	Hospitalization or Prolonged Hospitalization	A0705	Battery Problem	22/06/2022	22/06/2022
				A070504	Premature Discharge of Battery	22/06/2022	22/06/2022
		F1905	Device Revision or Replacement	A0705	Battery Problem	22/06/2022	22/06/2022
				A070504	Premature Discharge of Battery	22/06/2022	22/06/2022
CT2407	Active implantable	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem	22/06/2022	22/06/2022
				A070504	Premature Discharge of Battery	22/06/2022	22/06/2022
		F08	Hospitalization or Prolonged Hospitalization	A0705	Battery Problem	22/06/2022	22/06/2022
				A070504	Premature Discharge of Battery	22/06/2022	22/06/2022
		F1905	Device Revision or Replacement	A0705	Battery Problem	22/06/2022	22/06/2022
				A070504	Premature Discharge of Battery	22/06/2022	22/06/2022
CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem	22/06/2022	22/06/2022
				A070504	Premature Discharge of Battery	22/06/2022	22/06/2022
		F08	Hospitalization or Prolonged Hospitalization	A0705	Battery Problem	22/06/2022	22/06/2022
				A070504	Premature Discharge of Battery	22/06/2022	22/06/2022
		F1905	Device Revision or Replacement	A0705	Battery Problem	22/06/2022	22/06/2022
				A070504	Premature Discharge of Battery	22/06/2022	22/06/2022
CT2727	Active implant devices/ancillaries	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem	22/06/2022	22/06/2022
				A070504	Premature Discharge of Battery	22/06/2022	22/06/2022
		F08	Hospitalization or Prolonged Hospitalization	A0705	Battery Problem	22/06/2022	22/06/2022
				A070504	Premature Discharge of Battery	22/06/2022	22/06/2022
		F1905	Device Revision or Replacement	A0705	Battery Problem	22/06/2022	22/06/2022
				A070504	Premature Discharge of Battery	22/06/2022	22/06/2022
CT2750	Implantable cardiac electrical stimulators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem	22/06/2022	22/06/2022
				A070504	Premature Discharge of Battery	22/06/2022	22/06/2022
		F08	Hospitalization or Prolonged Hospitalization	A0705	Battery Problem	22/06/2022	22/06/2022
				A070504	Premature Discharge of Battery	22/06/2022	22/06/2022
		F1905	Device Revision or Replacement	A0705	Battery Problem	22/06/2022	22/06/2022
				A070504	Premature Discharge of Battery	22/06/2022	22/06/2022
CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem	22/06/2022	22/06/2022
				A070504	Premature Discharge of Battery	22/06/2022	22/06/2022
		F08	Hospitalization or Prolonged Hospitalization	A0705	Battery Problem	22/06/2022	22/06/2022
				A070504	Premature Discharge of Battery	22/06/2022	22/06/2022
		F1905	Device Revision or Replacement	A0705	Battery Problem	22/06/2022	22/06/2022
				A070504	Premature Discharge of Battery	22/06/2022	22/06/2022
CT2985	Active	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem	22/06/2022	22/06/2022
				A070504	Premature Discharge of Battery	22/06/2022	22/06/2022
		F08	Hospitalization or Prolonged Hospitalization	A0705	Battery Problem	22/06/2022	22/06/2022
				A070504	Premature Discharge of Battery	22/06/2022	22/06/2022
		F1905	Device Revision or Replacement	A0705	Battery Problem	22/06/2022	22/06/2022
				A070504	Premature Discharge of Battery	22/06/2022	22/06/2022
CT3001	Implantable metal	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem	22/06/2022	22/06/2022
				A070504	Premature Discharge of Battery	22/06/2022	22/06/2022
		F08	Hospitalization or Prolonged Hospitalization	A0705	Battery Problem	22/06/2022	22/06/2022
				A070504	Premature Discharge of Battery	22/06/2022	22/06/2022
		F1905	Device Revision or Replacement	A0705	Battery Problem	22/06/2022	22/06/2022
				A070504	Premature Discharge of Battery	22/06/2022	22/06/2022
CT3002	Implantable synthetic polymer	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem	22/06/2022	22/06/2022
				A070504	Premature Discharge of Battery	22/06/2022	22/06/2022
		F08	Hospitalization or Prolonged Hospitalization	A0705	Battery Problem	22/06/2022	22/06/2022
				A070504	Premature Discharge of Battery	22/06/2022	22/06/2022
		F1905	Device Revision or Replacement	A0705	Battery Problem	22/06/2022	22/06/2022
				A070504	Premature Discharge of Battery	22/06/2022	22/06/2022
CT3006	Implantable inorganic material	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem	22/06/2022	22/06/2022
				A070504	Premature Discharge of Battery	22/06/2022	22/06/2022
		F08	Hospitalization or Prolonged Hospitalization	A0705	Battery Problem	22/06/2022	22/06/2022
				A070504	Premature Discharge of Battery	22/06/2022	22/06/2022
		F1905	Device Revision or Replacement	A0705	Battery Problem	22/06/2022	22/06/2022
				A070504	Premature Discharge of Battery	22/06/2022	22/06/2022
CT314	Electrically powered	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem	22/06/2022	22/06/2022
				A070504	Premature Discharge of Battery	22/06/2022	22/06/2022
		F08	Hospitalization or Prolonged Hospitalization	A0705	Battery Problem	22/06/2022	22/06/2022
				A070504	Premature Discharge of Battery	22/06/2022	22/06/2022
		F1905	Device Revision or Replacement	A0705	Battery Problem	22/06/2022	22/06/2022
				A070504	Premature Discharge of Battery	22/06/2022	22/06/2022

					CT321	Long-term surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem	22/06/2022	22/06/2022
									A070504	Premature Discharge of Battery	22/06/2022	22/06/2022
							F08	Hospitalization or Prolonged Hospitalization	A0705	Battery Problem	22/06/2022	22/06/2022
									A070504	Premature Discharge of Battery	22/06/2022	22/06/2022
							F1905	Device Revision or Replacement	A0705	Battery Problem	22/06/2022	22/06/2022
									A070504	Premature Discharge of Battery	22/06/2022	22/06/2022
					CT449	Generators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem	22/06/2022	22/06/2022
									A070504	Premature Discharge of Battery	22/06/2022	22/06/2022
							F08	Hospitalization or Prolonged Hospitalization	A0705	Battery Problem	22/06/2022	22/06/2022
									A070504	Premature Discharge of Battery	22/06/2022	22/06/2022
							F1905	Device Revision or Replacement	A0705	Battery Problem	22/06/2022	22/06/2022
									A070504	Premature Discharge of Battery	22/06/2022	22/06/2022
					CT981	Single use	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem	22/06/2022	22/06/2022
									A070504	Premature Discharge of Battery	22/06/2022	22/06/2022
							F08	Hospitalization or Prolonged Hospitalization	A0705	Battery Problem	22/06/2022	22/06/2022
									A070504	Premature Discharge of Battery	22/06/2022	22/06/2022
							F1905	Device Revision or Replacement	A0705	Battery Problem	22/06/2022	22/06/2022
									A070504	Premature Discharge of Battery	22/06/2022	22/06/2022
					CT983	Surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem	22/06/2022	22/06/2022
									A070504	Premature Discharge of Battery	22/06/2022	22/06/2022
							F08	Hospitalization or Prolonged Hospitalization	A0705	Battery Problem	22/06/2022	22/06/2022
									A070504	Premature Discharge of Battery	22/06/2022	22/06/2022
							F1905	Device Revision or Replacement	A0705	Battery Problem	22/06/2022	22/06/2022
									A070504	Premature Discharge of Battery	22/06/2022	22/06/2022
6472	2022	All other reportable incidents	47265	Dual-chamber implantable pacemaker, rate-responsive	CT141	Electrodes/Leads and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
					CT145	Cardiovascular devices	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
					CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
					CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
					CT233	Surgical	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
					CT2406	Implantable	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
					CT2407	Active implantable	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
					CT2727	Active implant devices/ancillaries	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
					CT2750	Implantable cardiac electrical stimulators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
					CT2985	Active	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
					CT3001	Implantable metal	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
					CT3002	Implantable synthetic polymer	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
					CT3006	Implantable inorganic material	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
					CT314	Electrically powered	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		

					CT321	Long-term surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
					CT449	Generators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
					CT981	Single use	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
					CT983	Surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
6473	2022	All other reportable incidents	47265	Dual-chamber implantable pacemaker, rate-responsive	CT141	Electrodes/Leads and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem	13/04/2022	13/04/2022
									A070504	Premature Discharge of Battery	13/04/2022	13/04/2022
							F08	Hospitalization or Prolonged Hospitalization	A0705	Battery Problem	13/04/2022	13/04/2022
									A070504	Premature Discharge of Battery	13/04/2022	13/04/2022
							F1905	Device Revision or Replacement	A0705	Battery Problem	13/04/2022	13/04/2022
									A070504	Premature Discharge of Battery	13/04/2022	13/04/2022
					CT145	Cardiovascular devices	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem	13/04/2022	13/04/2022
									A070504	Premature Discharge of Battery	13/04/2022	13/04/2022
							F08	Hospitalization or Prolonged Hospitalization	A0705	Battery Problem	13/04/2022	13/04/2022
									A070504	Premature Discharge of Battery	13/04/2022	13/04/2022
							F1905	Device Revision or Replacement	A0705	Battery Problem	13/04/2022	13/04/2022
									A070504	Premature Discharge of Battery	13/04/2022	13/04/2022
					CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem	13/04/2022	13/04/2022
									A070504	Premature Discharge of Battery	13/04/2022	13/04/2022
							F08	Hospitalization or Prolonged Hospitalization	A0705	Battery Problem	13/04/2022	13/04/2022
									A070504	Premature Discharge of Battery	13/04/2022	13/04/2022
							F1905	Device Revision or Replacement	A0705	Battery Problem	13/04/2022	13/04/2022
									A070504	Premature Discharge of Battery	13/04/2022	13/04/2022
					CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem	13/04/2022	13/04/2022
									A070504	Premature Discharge of Battery	13/04/2022	13/04/2022
							F08	Hospitalization or Prolonged Hospitalization	A0705	Battery Problem	13/04/2022	13/04/2022
									A070504	Premature Discharge of Battery	13/04/2022	13/04/2022
							F1905	Device Revision or Replacement	A0705	Battery Problem	13/04/2022	13/04/2022
									A070504	Premature Discharge of Battery	13/04/2022	13/04/2022
					CT233	Surgical	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem	13/04/2022	13/04/2022
									A070504	Premature Discharge of Battery	13/04/2022	13/04/2022
							F08	Hospitalization or Prolonged Hospitalization	A0705	Battery Problem	13/04/2022	13/04/2022
									A070504	Premature Discharge of Battery	13/04/2022	13/04/2022
							F1905	Device Revision or Replacement	A0705	Battery Problem	13/04/2022	13/04/2022
									A070504	Premature Discharge of Battery	13/04/2022	13/04/2022
					CT2406	Implantable	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem	13/04/2022	13/04/2022
									A070504	Premature Discharge of Battery	13/04/2022	13/04/2022
							F08	Hospitalization or Prolonged Hospitalization	A0705	Battery Problem	13/04/2022	13/04/2022
									A070504	Premature Discharge of Battery	13/04/2022	13/04/2022
							F1905	Device Revision or Replacement	A0705	Battery Problem	13/04/2022	13/04/2022
									A070504	Premature Discharge of Battery	13/04/2022	13/04/2022
					CT2407	Active implantable	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem	13/04/2022	13/04/2022
									A070504	Premature Discharge of Battery	13/04/2022	13/04/2022
							F08	Hospitalization or Prolonged Hospitalization	A0705	Battery Problem	13/04/2022	13/04/2022
									A070504	Premature Discharge of Battery	13/04/2022	13/04/2022
							F1905	Device Revision or Replacement	A0705	Battery Problem	13/04/2022	13/04/2022
									A070504	Premature Discharge of Battery	13/04/2022	13/04/2022
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem	13/04/2022	13/04/2022
									A070504	Premature Discharge of Battery	13/04/2022	13/04/2022
							F08	Hospitalization or Prolonged Hospitalization	A0705	Battery Problem	13/04/2022	13/04/2022
									A070504	Premature Discharge of Battery	13/04/2022	13/04/2022
							F1905	Device Revision or Replacement	A0705	Battery Problem	13/04/2022	13/04/2022
									A070504	Premature Discharge of Battery	13/04/2022	13/04/2022
					CT2727	Active implant devices/ancillaries	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem	13/04/2022	13/04/2022
									A070504	Premature Discharge of Battery	13/04/2022	13/04/2022
							F08	Hospitalization or Prolonged Hospitalization	A0705	Battery Problem	13/04/2022	13/04/2022
									A070504	Premature Discharge of Battery	13/04/2022	13/04/2022
							F1905	Device Revision or Replacement	A0705	Battery Problem	13/04/2022	13/04/2022
									A070504	Premature Discharge of Battery	13/04/2022	13/04/2022
					CT2750	Implantable cardiac electrical stimulators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem	13/04/2022	13/04/2022
									A070504	Premature Discharge of Battery	13/04/2022	13/04/2022
							F08	Hospitalization or Prolonged Hospitalization	A0705	Battery Problem	13/04/2022	13/04/2022
									A070504	Premature Discharge of Battery	13/04/2022	13/04/2022
							F1905	Device Revision or Replacement	A0705	Battery Problem	13/04/2022	13/04/2022
									A070504	Premature Discharge of Battery	13/04/2022	13/04/2022
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem	13/04/2022	13/04/2022
									A070504	Premature Discharge of Battery	13/04/2022	13/04/2022
							F08	Hospitalization or Prolonged Hospitalization	A0705	Battery Problem	13/04/2022	13/04/2022
									A070504	Premature Discharge of Battery	13/04/2022	13/04/2022
							F1905	Device Revision or Replacement	A0705	Battery Problem	13/04/2022	13/04/2022
									A070504	Premature Discharge of Battery	13/04/2022	13/04/2022
					CT2985	Active	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem	13/04/2022	13/04/2022
									A070504	Premature Discharge of Battery	13/04/2022	13/04/2022
							F08	Hospitalization or Prolonged Hospitalization	A0705	Battery Problem	13/04/2022	13/04/2022
									A070504	Premature Discharge of Battery	13/04/2022	13/04/2022
							F1905	Device Revision or Replacement	A0705	Battery Problem	13/04/2022	13/04/2022
									A070504	Premature Discharge of Battery	13/04/2022	13/04/2022

					CT3001	Implantable metal	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem	13/04/2022	13/04/2022
									A070504	Premature Discharge of Battery	13/04/2022	13/04/2022
							F08	Hospitalization or Prolonged Hospitalization	A0705	Battery Problem	13/04/2022	13/04/2022
									A070504	Premature Discharge of Battery	13/04/2022	13/04/2022
							F1905	Device Revision or Replacement	A0705	Battery Problem	13/04/2022	13/04/2022
									A070504	Premature Discharge of Battery	13/04/2022	13/04/2022
					CT3002	Implantable synthetic polymer	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem	13/04/2022	13/04/2022
									A070504	Premature Discharge of Battery	13/04/2022	13/04/2022
							F08	Hospitalization or Prolonged Hospitalization	A0705	Battery Problem	13/04/2022	13/04/2022
									A070504	Premature Discharge of Battery	13/04/2022	13/04/2022
							F1905	Device Revision or Replacement	A0705	Battery Problem	13/04/2022	13/04/2022
									A070504	Premature Discharge of Battery	13/04/2022	13/04/2022
					CT3006	Implantable inorganic material	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem	13/04/2022	13/04/2022
									A070504	Premature Discharge of Battery	13/04/2022	13/04/2022
							F08	Hospitalization or Prolonged Hospitalization	A0705	Battery Problem	13/04/2022	13/04/2022
									A070504	Premature Discharge of Battery	13/04/2022	13/04/2022
							F1905	Device Revision or Replacement	A0705	Battery Problem	13/04/2022	13/04/2022
									A070504	Premature Discharge of Battery	13/04/2022	13/04/2022
					CT314	Electrically powered	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem	13/04/2022	13/04/2022
									A070504	Premature Discharge of Battery	13/04/2022	13/04/2022
							F08	Hospitalization or Prolonged Hospitalization	A0705	Battery Problem	13/04/2022	13/04/2022
									A070504	Premature Discharge of Battery	13/04/2022	13/04/2022
							F1905	Device Revision or Replacement	A0705	Battery Problem	13/04/2022	13/04/2022
									A070504	Premature Discharge of Battery	13/04/2022	13/04/2022
					CT321	Long-term surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem	13/04/2022	13/04/2022
									A070504	Premature Discharge of Battery	13/04/2022	13/04/2022
							F08	Hospitalization or Prolonged Hospitalization	A0705	Battery Problem	13/04/2022	13/04/2022
									A070504	Premature Discharge of Battery	13/04/2022	13/04/2022
							F1905	Device Revision or Replacement	A0705	Battery Problem	13/04/2022	13/04/2022
									A070504	Premature Discharge of Battery	13/04/2022	13/04/2022
					CT449	Generators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem	13/04/2022	13/04/2022
									A070504	Premature Discharge of Battery	13/04/2022	13/04/2022
							F08	Hospitalization or Prolonged Hospitalization	A0705	Battery Problem	13/04/2022	13/04/2022
									A070504	Premature Discharge of Battery	13/04/2022	13/04/2022
							F1905	Device Revision or Replacement	A0705	Battery Problem	13/04/2022	13/04/2022
									A070504	Premature Discharge of Battery	13/04/2022	13/04/2022
					CT981	Single use	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem	13/04/2022	13/04/2022
									A070504	Premature Discharge of Battery	13/04/2022	13/04/2022
							F08	Hospitalization or Prolonged Hospitalization	A0705	Battery Problem	13/04/2022	13/04/2022
									A070504	Premature Discharge of Battery	13/04/2022	13/04/2022
							F1905	Device Revision or Replacement	A0705	Battery Problem	13/04/2022	13/04/2022
									A070504	Premature Discharge of Battery	13/04/2022	13/04/2022
					CT983	Surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem	13/04/2022	13/04/2022
									A070504	Premature Discharge of Battery	13/04/2022	13/04/2022
							F08	Hospitalization or Prolonged Hospitalization	A0705	Battery Problem	13/04/2022	13/04/2022
									A070504	Premature Discharge of Battery	13/04/2022	13/04/2022
							F1905	Device Revision or Replacement	A0705	Battery Problem	13/04/2022	13/04/2022
									A070504	Premature Discharge of Battery	13/04/2022	13/04/2022
6474	2022	All other reportable incidents	47267	Single-chamber implantable pacemaker, rate-responsive	CT141	Electrodes/Leads and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	22/04/2022	22/04/2022
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	22/04/2022	22/04/2022
							F1903	Device Explantation	A070504	Premature Discharge of Battery	22/04/2022	22/04/2022
					CT145	Cardiovascular devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	22/04/2022	22/04/2022
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	22/04/2022	22/04/2022
							F1903	Device Explantation	A070504	Premature Discharge of Battery	22/04/2022	22/04/2022
					CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	22/04/2022	22/04/2022
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	22/04/2022	22/04/2022
							F1903	Device Explantation	A070504	Premature Discharge of Battery	22/04/2022	22/04/2022
					CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	22/04/2022	22/04/2022
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	22/04/2022	22/04/2022
							F1903	Device Explantation	A070504	Premature Discharge of Battery	22/04/2022	22/04/2022
					CT233	Surgical	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	22/04/2022	22/04/2022
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	22/04/2022	22/04/2022
							F1903	Device Explantation	A070504	Premature Discharge of Battery	22/04/2022	22/04/2022
					CT2406	Implantable	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	22/04/2022	22/04/2022
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	22/04/2022	22/04/2022
							F1903	Device Explantation	A070504	Premature Discharge of Battery	22/04/2022	22/04/2022
					CT2407	Active implantable	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	22/04/2022	22/04/2022
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	22/04/2022	22/04/2022
							F1903	Device Explantation	A070504	Premature Discharge of Battery	22/04/2022	22/04/2022
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	22/04/2022	22/04/2022
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	22/04/2022	22/04/2022
							F1903	Device Explantation	A070504	Premature Discharge of Battery	22/04/2022	22/04/2022
					CT2727	Active implant devices/ancillaries	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	22/04/2022	22/04/2022
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	22/04/2022	22/04/2022
							F1903	Device Explantation	A070504	Premature Discharge of Battery	22/04/2022	22/04/2022
					CT2750	Implantable cardiac electrical stimulators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	22/04/2022	22/04/2022
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	22/04/2022	22/04/2022
							F1903	Device Explantation	A070504	Premature Discharge of Battery	22/04/2022	22/04/2022
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	22/04/2022	22/04/2022
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	22/04/2022	22/04/2022
							F1903	Device Explantation	A070504	Premature Discharge of Battery	22/04/2022	22/04/2022
					CT2985	Active	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	22/04/2022	22/04/2022
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	22/04/2022	22/04/2022
							F1903	Device Explantation	A070504	Premature Discharge of Battery	22/04/2022	22/04/2022
					CT3001	Implantable metal	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	22/04/2022	22/04/2022
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	22/04/2022	22/04/2022
							F1903	Device Explantation	A070504	Premature Discharge of Battery	22/04/2022	22/04/2022
					CT3002	Implantable synthetic polymer	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	22/04/2022	22/04/2022

							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	22/04/2022	22/04/2022
							F1903	Device Explantation	A070504	Premature Discharge of Battery	22/04/2022	22/04/2022
					CT3006	Implantable inorganic material	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	22/04/2022	22/04/2022
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	22/04/2022	22/04/2022
							F1903	Device Explantation	A070504	Premature Discharge of Battery	22/04/2022	22/04/2022
					CT314	Electrically powered	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	22/04/2022	22/04/2022
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	22/04/2022	22/04/2022
							F1903	Device Explantation	A070504	Premature Discharge of Battery	22/04/2022	22/04/2022
					CT321	Long-term surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	22/04/2022	22/04/2022
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	22/04/2022	22/04/2022
							F1903	Device Explantation	A070504	Premature Discharge of Battery	22/04/2022	22/04/2022
					CT449	Generators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	22/04/2022	22/04/2022
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	22/04/2022	22/04/2022
							F1903	Device Explantation	A070504	Premature Discharge of Battery	22/04/2022	22/04/2022
					CT981	Single use	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	22/04/2022	22/04/2022
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	22/04/2022	22/04/2022
							F1903	Device Explantation	A070504	Premature Discharge of Battery	22/04/2022	22/04/2022
					CT983	Surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	22/04/2022	22/04/2022
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	22/04/2022	22/04/2022
							F1903	Device Explantation	A070504	Premature Discharge of Battery	22/04/2022	22/04/2022
6475	2022	All other reportable incidents	36241	Pacemaker/defibrillator lead	CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070909	Oversensing		
									A072201	High Impedance		
							F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A070909	Oversensing		
									A072201	High Impedance		
6476	2022	All other reportable incidents	47265	Dual-chamber implantable pacemaker, rate-responsive	CT141	Electrodes/Leads and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem	28/07/2022	28/07/2022
									A070504	Premature Discharge of Battery	28/07/2022	28/07/2022
							F08	Hospitalization or Prolonged Hospitalization	A0705	Battery Problem	28/07/2022	28/07/2022
									A070504	Premature Discharge of Battery	28/07/2022	28/07/2022
							F1905	Device Revision or Replacement	A0705	Battery Problem	28/07/2022	28/07/2022
									A070504	Premature Discharge of Battery	28/07/2022	28/07/2022
					CT145	Cardiovascular devices	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem	28/07/2022	28/07/2022
									A070504	Premature Discharge of Battery	28/07/2022	28/07/2022
							F08	Hospitalization or Prolonged Hospitalization	A0705	Battery Problem	28/07/2022	28/07/2022
									A070504	Premature Discharge of Battery	28/07/2022	28/07/2022
							F1905	Device Revision or Replacement	A0705	Battery Problem	28/07/2022	28/07/2022
									A070504	Premature Discharge of Battery	28/07/2022	28/07/2022
					CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem	28/07/2022	28/07/2022
									A070504	Premature Discharge of Battery	28/07/2022	28/07/2022
							F08	Hospitalization or Prolonged Hospitalization	A0705	Battery Problem	28/07/2022	28/07/2022
									A070504	Premature Discharge of Battery	28/07/2022	28/07/2022
							F1905	Device Revision or Replacement	A0705	Battery Problem	28/07/2022	28/07/2022
									A070504	Premature Discharge of Battery	28/07/2022	28/07/2022
					CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem	28/07/2022	28/07/2022
									A070504	Premature Discharge of Battery	28/07/2022	28/07/2022
							F08	Hospitalization or Prolonged Hospitalization	A0705	Battery Problem	28/07/2022	28/07/2022
									A070504	Premature Discharge of Battery	28/07/2022	28/07/2022
							F1905	Device Revision or Replacement	A0705	Battery Problem	28/07/2022	28/07/2022
									A070504	Premature Discharge of Battery	28/07/2022	28/07/2022
					CT233	Surgical	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem	28/07/2022	28/07/2022
									A070504	Premature Discharge of Battery	28/07/2022	28/07/2022
							F08	Hospitalization or Prolonged Hospitalization	A0705	Battery Problem	28/07/2022	28/07/2022
									A070504	Premature Discharge of Battery	28/07/2022	28/07/2022
							F1905	Device Revision or Replacement	A0705	Battery Problem	28/07/2022	28/07/2022
									A070504	Premature Discharge of Battery	28/07/2022	28/07/2022
					CT2406	Implantable	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem	28/07/2022	28/07/2022
									A070504	Premature Discharge of Battery	28/07/2022	28/07/2022
							F08	Hospitalization or Prolonged Hospitalization	A0705	Battery Problem	28/07/2022	28/07/2022
									A070504	Premature Discharge of Battery	28/07/2022	28/07/2022
							F1905	Device Revision or Replacement	A0705	Battery Problem	28/07/2022	28/07/2022
									A070504	Premature Discharge of Battery	28/07/2022	28/07/2022
					CT2407	Active implantable	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem	28/07/2022	28/07/2022
									A070504	Premature Discharge of Battery	28/07/2022	28/07/2022
							F08	Hospitalization or Prolonged Hospitalization	A0705	Battery Problem	28/07/2022	28/07/2022
									A070504	Premature Discharge of Battery	28/07/2022	28/07/2022
							F1905	Device Revision or Replacement	A0705	Battery Problem	28/07/2022	28/07/2022
									A070504	Premature Discharge of Battery	28/07/2022	28/07/2022
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem	28/07/2022	28/07/2022
									A070504	Premature Discharge of Battery	28/07/2022	28/07/2022
							F08	Hospitalization or Prolonged Hospitalization	A0705	Battery Problem	28/07/2022	28/07/2022
									A070504	Premature Discharge of Battery	28/07/2022	28/07/2022
							F1905	Device Revision or Replacement	A0705	Battery Problem	28/07/2022	28/07/2022
									A070504	Premature Discharge of Battery	28/07/2022	28/07/2022
					CT2727	Active implant devices/ancillaries	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem	28/07/2022	28/07/2022
									A070504	Premature Discharge of Battery	28/07/2022	28/07/2022
							F08	Hospitalization or Prolonged Hospitalization	A0705	Battery Problem	28/07/2022	28/07/2022
									A070504	Premature Discharge of Battery	28/07/2022	28/07/2022
							F1905	Device Revision or Replacement	A0705	Battery Problem	28/07/2022	28/07/2022
									A070504	Premature Discharge of Battery	28/07/2022	28/07/2022
					CT2750	Implantable cardiac electrical stimulators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem	28/07/2022	28/07/2022
									A070504	Premature Discharge of Battery	28/07/2022	28/07/2022
							F08	Hospitalization or Prolonged Hospitalization	A0705	Battery Problem	28/07/2022	28/07/2022
									A070504	Premature Discharge of Battery	28/07/2022	28/07/2022
							F1905	Device Revision or Replacement	A0705	Battery Problem	28/07/2022	28/07/2022
									A070504	Premature Discharge of Battery	28/07/2022	28/07/2022
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem	28/07/2022	28/07/2022
									A070504	Premature Discharge of Battery	28/07/2022	28/07/2022
							F08	Hospitalization or Prolonged Hospitalization	A0705	Battery Problem	28/07/2022	28/07/2022
									A070504	Premature Discharge of Battery	28/07/2022	28/07/2022

							F1905	Device Revision or Replacement	A0705	Battery Problem	28/07/2022	28/07/2022
									A070504	Premature Discharge of Battery	28/07/2022	28/07/2022
					CT2985	Active	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem	28/07/2022	28/07/2022
									A070504	Premature Discharge of Battery	28/07/2022	28/07/2022
							F08	Hospitalization or Prolonged Hospitalization	A0705	Battery Problem	28/07/2022	28/07/2022
									A070504	Premature Discharge of Battery	28/07/2022	28/07/2022
							F1905	Device Revision or Replacement	A0705	Battery Problem	28/07/2022	28/07/2022
									A070504	Premature Discharge of Battery	28/07/2022	28/07/2022
					CT3001	Implantable metal	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem	28/07/2022	28/07/2022
									A070504	Premature Discharge of Battery	28/07/2022	28/07/2022
							F08	Hospitalization or Prolonged Hospitalization	A0705	Battery Problem	28/07/2022	28/07/2022
									A070504	Premature Discharge of Battery	28/07/2022	28/07/2022
							F1905	Device Revision or Replacement	A0705	Battery Problem	28/07/2022	28/07/2022
									A070504	Premature Discharge of Battery	28/07/2022	28/07/2022
					CT3002	Implantable synthetic polymer	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem	28/07/2022	28/07/2022
									A070504	Premature Discharge of Battery	28/07/2022	28/07/2022
							F08	Hospitalization or Prolonged Hospitalization	A0705	Battery Problem	28/07/2022	28/07/2022
									A070504	Premature Discharge of Battery	28/07/2022	28/07/2022
							F1905	Device Revision or Replacement	A0705	Battery Problem	28/07/2022	28/07/2022
									A070504	Premature Discharge of Battery	28/07/2022	28/07/2022
					CT3006	Implantable inorganic material	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem	28/07/2022	28/07/2022
									A070504	Premature Discharge of Battery	28/07/2022	28/07/2022
							F08	Hospitalization or Prolonged Hospitalization	A0705	Battery Problem	28/07/2022	28/07/2022
									A070504	Premature Discharge of Battery	28/07/2022	28/07/2022
							F1905	Device Revision or Replacement	A0705	Battery Problem	28/07/2022	28/07/2022
									A070504	Premature Discharge of Battery	28/07/2022	28/07/2022
					CT314	Electrically powered	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem	28/07/2022	28/07/2022
									A070504	Premature Discharge of Battery	28/07/2022	28/07/2022
							F08	Hospitalization or Prolonged Hospitalization	A0705	Battery Problem	28/07/2022	28/07/2022
									A070504	Premature Discharge of Battery	28/07/2022	28/07/2022
							F1905	Device Revision or Replacement	A0705	Battery Problem	28/07/2022	28/07/2022
									A070504	Premature Discharge of Battery	28/07/2022	28/07/2022
					CT321	Long-term surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem	28/07/2022	28/07/2022
									A070504	Premature Discharge of Battery	28/07/2022	28/07/2022
							F08	Hospitalization or Prolonged Hospitalization	A0705	Battery Problem	28/07/2022	28/07/2022
									A070504	Premature Discharge of Battery	28/07/2022	28/07/2022
							F1905	Device Revision or Replacement	A0705	Battery Problem	28/07/2022	28/07/2022
									A070504	Premature Discharge of Battery	28/07/2022	28/07/2022
					CT449	Generators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem	28/07/2022	28/07/2022
									A070504	Premature Discharge of Battery	28/07/2022	28/07/2022
							F08	Hospitalization or Prolonged Hospitalization	A0705	Battery Problem	28/07/2022	28/07/2022
									A070504	Premature Discharge of Battery	28/07/2022	28/07/2022
							F1905	Device Revision or Replacement	A0705	Battery Problem	28/07/2022	28/07/2022
									A070504	Premature Discharge of Battery	28/07/2022	28/07/2022
					CT981	Single use	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem	28/07/2022	28/07/2022
									A070504	Premature Discharge of Battery	28/07/2022	28/07/2022
							F08	Hospitalization or Prolonged Hospitalization	A0705	Battery Problem	28/07/2022	28/07/2022
									A070504	Premature Discharge of Battery	28/07/2022	28/07/2022
							F1905	Device Revision or Replacement	A0705	Battery Problem	28/07/2022	28/07/2022
									A070504	Premature Discharge of Battery	28/07/2022	28/07/2022
					CT983	Surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem	28/07/2022	28/07/2022
									A070504	Premature Discharge of Battery	28/07/2022	28/07/2022
							F08	Hospitalization or Prolonged Hospitalization	A0705	Battery Problem	28/07/2022	28/07/2022
									A070504	Premature Discharge of Battery	28/07/2022	28/07/2022
							F1905	Device Revision or Replacement	A0705	Battery Problem	28/07/2022	28/07/2022
									A070504	Premature Discharge of Battery	28/07/2022	28/07/2022
6477	2022	All other reportable incidents	37265	Dual-chamber implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A070910	Undersensing		
									A1301	Failure to Read Input Signal		
							F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A070910	Undersensing		
									A1301	Failure to Read Input Signal		
6478	2021	All other reportable incidents	17882	Physiologic-monitoring defibrillation system	CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070908	Failure to Sense		
									A1107	Data Problem		
							F27	Problem Identified During Non-Clinical Procedure	A070908	Failure to Sense		
									A1107	Data Problem		
6479	2021	All other reportable incidents	35852	Single-chamber implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem	17/02/2022	17/02/2022
									A070504	Premature Discharge of Battery	17/02/2022	17/02/2022
							F08	Hospitalization or Prolonged Hospitalization	A0705	Battery Problem	17/02/2022	17/02/2022
									A070504	Premature Discharge of Battery	17/02/2022	17/02/2022
							F1905	Device Revision or Replacement	A0705	Battery Problem	17/02/2022	17/02/2022
									A070504	Premature Discharge of Battery	17/02/2022	17/02/2022
6480	2021	All other reportable incidents	35223	Endocardial/interventricular septal pacing lead	CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070901	Decreased Sensitivity		
									A070909	Oversensing		
									A090801	Signal Artifact/Noise		
							F26	No Health Consequences or Impact	A070901	Decreased Sensitivity		
									A070909	Oversensing		
									A090801	Signal Artifact/Noise		
6481	2021	All other reportable incidents	35223	Endocardial/interventricular septal pacing lead	CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070909	Oversensing	24/08/2021	24/08/2021
							F1901	Additional Surgery	A070909	Oversensing	24/08/2021	24/08/2021
6482	2021	All other reportable incidents	35223	Endocardial/interventricular septal pacing lead	CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A0701	Capturing Problem	24/08/2021	24/08/2021
									A070909	Oversensing	24/08/2021	24/08/2021
							F1901	Additional Surgery	A0701	Capturing Problem	24/08/2021	24/08/2021
									A070909	Oversensing	24/08/2021	24/08/2021
6483	2022	All other reportable incidents	47270	Cardiac resynchronization therapy implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E0601	Arrhythmia	A070101	Failure to Capture	18/04/2022	18/04/2022
									A070504	Premature Discharge of Battery	18/04/2022	18/04/2022
									A071102	Failure to Interrogate	18/04/2022	18/04/2022
									A071204	No Pacing	18/04/2022	18/04/2022
									A160503	Inappropriate or Unexpected Reset	18/04/2022	18/04/2022
							E2403	No Clinical Signs, Symptoms or Conditions	A070101	Failure to Capture	18/04/2022	18/04/2022

									A070504	Premature Discharge of Battery	18/04/2022	18/04/2022
									A071102	Failure to Interrogate	18/04/2022	18/04/2022
									A071204	No Pacing	18/04/2022	18/04/2022
									A160503	Inappropriate or Unexpected Reset	18/04/2022	18/04/2022
							F1901	Additional Surgery	A070101	Failure to Capture	18/04/2022	18/04/2022
									A070504	Premature Discharge of Battery	18/04/2022	18/04/2022
									A071102	Failure to Interrogate	18/04/2022	18/04/2022
									A071204	No Pacing	18/04/2022	18/04/2022
									A160503	Inappropriate or Unexpected Reset	18/04/2022	18/04/2022
6484	2021	All other reportable incidents	17882	Physiologic-monitoring defibrillation system	CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A0719	Unexpected Shutdown		
							F26	No Health Consequences or Impact	A0719	Unexpected Shutdown		
6485	2021	All other reportable incidents	17882	Physiologic-monitoring defibrillation system	CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070912	Incorrect Interpretation of Signal		
							F26	No Health Consequences or Impact	A070912	Incorrect Interpretation of Signal		
6486	2021	All other reportable incidents	17882	Physiologic-monitoring defibrillation system	CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070912	Incorrect Interpretation of Signal		
							F26	No Health Consequences or Impact	A070912	Incorrect Interpretation of Signal		
6487	2021	All other reportable incidents	47270	Cardiac resynchronization therapy implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A0908	Incorrect, Inadequate or Imprecise Result or Readings		
							F26	No Health Consequences or Impact	A0908	Incorrect, Inadequate or Imprecise Result or Readings		
6488	2021	All other reportable incidents	37265	Dual-chamber implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
									A072201	High Impedance		
							F26	No Health Consequences or Impact	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
									A072201	High Impedance		
6489	2021	All other reportable incidents	47270	Cardiac resynchronization therapy implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A070909	Oversensing		
									A0712	Pacing Problem		
							F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A070909	Oversensing		
									A0712	Pacing Problem		
6490	2021	All other reportable incidents	35852	Single-chamber implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E2104	Electric Shock	A070909	Oversensing		
									A071302	Inappropriate/Inadequate Shock/Stimulation		
							E2403	No Clinical Signs, Symptoms or Conditions	A070909	Oversensing		
									A071302	Inappropriate/Inadequate Shock/Stimulation		
							F26	No Health Consequences or Impact	A070909	Oversensing		
									A071302	Inappropriate/Inadequate Shock/Stimulation		
6491	2021	All other reportable incidents	36241	Pacemaker/defibrillator lead	CT1684	Cardiac pacemakers and associated devices	E2104	Electric Shock	A070909	Oversensing		
									A071302	Inappropriate/Inadequate Shock/Stimulation		
							E2403	No Clinical Signs, Symptoms or Conditions	A070909	Oversensing		
									A071302	Inappropriate/Inadequate Shock/Stimulation		
							F26	No Health Consequences or Impact	A070909	Oversensing		
									A071302	Inappropriate/Inadequate Shock/Stimulation		
6492	2021	All other reportable incidents	36241	Pacemaker/defibrillator lead	CT1684	Cardiac pacemakers and associated devices	E2104	Electric Shock	A070909	Oversensing		
									A071302	Inappropriate/Inadequate Shock/Stimulation		
							E2403	No Clinical Signs, Symptoms or Conditions	A070909	Oversensing		
									A071302	Inappropriate/Inadequate Shock/Stimulation		
							F26	No Health Consequences or Impact	A070909	Oversensing		
									A071302	Inappropriate/Inadequate Shock/Stimulation		
6493	2021	All other reportable incidents	35852	Single-chamber implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery		
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery		
6494	2021	All other reportable incidents	35223	Endocardial/interventricular septal pacing lead	CT1684	Cardiac pacemakers and associated devices	E011903	Syncope/Fainting	A0401	Break		
									A070909	Oversensing		
							E2403	No Clinical Signs, Symptoms or Conditions	A0401	Break		
									A070909	Oversensing		
							F22	Unexpected Diagnostic Intervention	A0401	Break		
									A070909	Oversensing		
6495	2021	All other reportable incidents	47263	Cardiac resynchronization therapy implantable pacemaker	CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F27	Problem Identified During Non-Clinical Procedure	A070504	Premature Discharge of Battery		
					CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F27	Problem Identified During Non-Clinical Procedure	A070504	Premature Discharge of Battery		
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F27	Problem Identified During Non-Clinical Procedure	A070504	Premature Discharge of Battery		
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F27	Problem Identified During Non-Clinical Procedure	A070504	Premature Discharge of Battery		
6496	2022	All other reportable incidents	35223	Endocardial/interventricular septal pacing lead	CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070909	Oversensing		
							F1905	Device Revision or Replacement	A070909	Oversensing		
6497	2022	All other reportable incidents	47265	Dual-chamber implantable pacemaker, rate-responsive	CT141	Electrodes/Leads and associated devices	E011903	Syncope/Fainting	A070504	Premature Discharge of Battery	01/04/2022	30/04/2022
							E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	01/04/2022	30/04/2022
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	01/04/2022	30/04/2022
							F1203	Life Threatening Illness or Injury	A070504	Premature Discharge of Battery	01/04/2022	30/04/2022
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	01/04/2022	30/04/2022
							F23	Unexpected Medical Intervention	A070504	Premature Discharge of Battery	01/04/2022	30/04/2022
					CT145	Cardiovascular devices	E011903	Syncope/Fainting	A070504	Premature Discharge of Battery	01/04/2022	30/04/2022
							E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	01/04/2022	30/04/2022
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	01/04/2022	30/04/2022
							F1203	Life Threatening Illness or Injury	A070504	Premature Discharge of Battery	01/04/2022	30/04/2022
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	01/04/2022	30/04/2022
							F23	Unexpected Medical Intervention	A070504	Premature Discharge of Battery	01/04/2022	30/04/2022
					CT1684	Cardiac pacemakers and associated devices	E011903	Syncope/Fainting	A070504	Premature Discharge of Battery	01/04/2022	30/04/2022
							E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	01/04/2022	30/04/2022
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	01/04/2022	30/04/2022
							F1203	Life Threatening Illness or Injury	A070504	Premature Discharge of Battery	01/04/2022	30/04/2022
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	01/04/2022	30/04/2022
							F23	Unexpected Medical Intervention	A070504	Premature Discharge of Battery	01/04/2022	30/04/2022
					CT1685	Cardiac pacemakers	E011903	Syncope/Fainting	A070504	Premature Discharge of Battery	01/04/2022	30/04/2022
							E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	01/04/2022	30/04/2022
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	01/04/2022	30/04/2022
							F1203	Life Threatening Illness or Injury	A070504	Premature Discharge of Battery	01/04/2022	30/04/2022

							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	01/04/2022	30/04/2022
							F23	Unexpected Medical Intervention	A070504	Premature Discharge of Battery	01/04/2022	30/04/2022
					CT233	Surgical	E011903	Syncope/Fainting	A070504	Premature Discharge of Battery	01/04/2022	30/04/2022
							E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	01/04/2022	30/04/2022
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	01/04/2022	30/04/2022
							F1203	Life Threatening Illness or Injury	A070504	Premature Discharge of Battery	01/04/2022	30/04/2022
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	01/04/2022	30/04/2022
							F23	Unexpected Medical Intervention	A070504	Premature Discharge of Battery	01/04/2022	30/04/2022
					CT2406	Implantable	E011903	Syncope/Fainting	A070504	Premature Discharge of Battery	01/04/2022	30/04/2022
							E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	01/04/2022	30/04/2022
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	01/04/2022	30/04/2022
							F1203	Life Threatening Illness or Injury	A070504	Premature Discharge of Battery	01/04/2022	30/04/2022
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	01/04/2022	30/04/2022
							F23	Unexpected Medical Intervention	A070504	Premature Discharge of Battery	01/04/2022	30/04/2022
					CT2407	Active implantable	E011903	Syncope/Fainting	A070504	Premature Discharge of Battery	01/04/2022	30/04/2022
							E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	01/04/2022	30/04/2022
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	01/04/2022	30/04/2022
							F1203	Life Threatening Illness or Injury	A070504	Premature Discharge of Battery	01/04/2022	30/04/2022
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	01/04/2022	30/04/2022
							F23	Unexpected Medical Intervention	A070504	Premature Discharge of Battery	01/04/2022	30/04/2022
					CT257	Implantable cardiac pacemakers	E011903	Syncope/Fainting	A070504	Premature Discharge of Battery	01/04/2022	30/04/2022
							E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	01/04/2022	30/04/2022
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	01/04/2022	30/04/2022
							F1203	Life Threatening Illness or Injury	A070504	Premature Discharge of Battery	01/04/2022	30/04/2022
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	01/04/2022	30/04/2022
							F23	Unexpected Medical Intervention	A070504	Premature Discharge of Battery	01/04/2022	30/04/2022
					CT2727	Active implant devices/ancillaries	E011903	Syncope/Fainting	A070504	Premature Discharge of Battery	01/04/2022	30/04/2022
							E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	01/04/2022	30/04/2022
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	01/04/2022	30/04/2022
							F1203	Life Threatening Illness or Injury	A070504	Premature Discharge of Battery	01/04/2022	30/04/2022
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	01/04/2022	30/04/2022
							F23	Unexpected Medical Intervention	A070504	Premature Discharge of Battery	01/04/2022	30/04/2022
					CT2750	Implantable cardiac electrical stimulators and associated devices	E011903	Syncope/Fainting	A070504	Premature Discharge of Battery	01/04/2022	30/04/2022
							E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	01/04/2022	30/04/2022
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	01/04/2022	30/04/2022
							F1203	Life Threatening Illness or Injury	A070504	Premature Discharge of Battery	01/04/2022	30/04/2022
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	01/04/2022	30/04/2022
							F23	Unexpected Medical Intervention	A070504	Premature Discharge of Battery	01/04/2022	30/04/2022
					CT2866	Implantable cardiac electrical stimulators	E011903	Syncope/Fainting	A070504	Premature Discharge of Battery	01/04/2022	30/04/2022
							E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	01/04/2022	30/04/2022
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	01/04/2022	30/04/2022
							F1203	Life Threatening Illness or Injury	A070504	Premature Discharge of Battery	01/04/2022	30/04/2022
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	01/04/2022	30/04/2022
							F23	Unexpected Medical Intervention	A070504	Premature Discharge of Battery	01/04/2022	30/04/2022
					CT2985	Active	E011903	Syncope/Fainting	A070504	Premature Discharge of Battery	01/04/2022	30/04/2022
							E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	01/04/2022	30/04/2022
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	01/04/2022	30/04/2022
							F1203	Life Threatening Illness or Injury	A070504	Premature Discharge of Battery	01/04/2022	30/04/2022
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	01/04/2022	30/04/2022
							F23	Unexpected Medical Intervention	A070504	Premature Discharge of Battery	01/04/2022	30/04/2022
					CT3001	Implantable metal	E011903	Syncope/Fainting	A070504	Premature Discharge of Battery	01/04/2022	30/04/2022
							E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	01/04/2022	30/04/2022
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	01/04/2022	30/04/2022
							F1203	Life Threatening Illness or Injury	A070504	Premature Discharge of Battery	01/04/2022	30/04/2022
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	01/04/2022	30/04/2022
							F23	Unexpected Medical Intervention	A070504	Premature Discharge of Battery	01/04/2022	30/04/2022
					CT3002	Implantable synthetic polymer	E011903	Syncope/Fainting	A070504	Premature Discharge of Battery	01/04/2022	30/04/2022
							E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	01/04/2022	30/04/2022
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	01/04/2022	30/04/2022
							F1203	Life Threatening Illness or Injury	A070504	Premature Discharge of Battery	01/04/2022	30/04/2022
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	01/04/2022	30/04/2022
							F23	Unexpected Medical Intervention	A070504	Premature Discharge of Battery	01/04/2022	30/04/2022
					CT3006	Implantable inorganic material	E011903	Syncope/Fainting	A070504	Premature Discharge of Battery	01/04/2022	30/04/2022
							E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	01/04/2022	30/04/2022
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	01/04/2022	30/04/2022
							F1203	Life Threatening Illness or Injury	A070504	Premature Discharge of Battery	01/04/2022	30/04/2022
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	01/04/2022	30/04/2022
							F23	Unexpected Medical Intervention	A070504	Premature Discharge of Battery	01/04/2022	30/04/2022
					CT314	Electrically powered	E011903	Syncope/Fainting	A070504	Premature Discharge of Battery	01/04/2022	30/04/2022
							E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	01/04/2022	30/04/2022
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	01/04/2022	30/04/2022
							F1203	Life Threatening Illness or Injury	A070504	Premature Discharge of Battery	01/04/2022	30/04/2022
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	01/04/2022	30/04/2022
							F23	Unexpected Medical Intervention	A070504	Premature Discharge of Battery	01/04/2022	30/04/2022
					CT321	Long-term surgical invasive	E011903	Syncope/Fainting	A070504	Premature Discharge of Battery	01/04/2022	30/04/2022
							E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	01/04/2022	30/04/2022
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	01/04/2022	30/04/2022
							F1203	Life Threatening Illness or Injury	A070504	Premature Discharge of Battery	01/04/2022	30/04/2022
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	01/04/2022	30/04/2022
							F23	Unexpected Medical Intervention	A070504	Premature Discharge of Battery	01/04/2022	30/04/2022
					CT449	Generators and associated devices	E011903	Syncope/Fainting	A070504	Premature Discharge of Battery	01/04/2022	30/04/2022
							E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	01/04/2022	30/04/2022
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	01/04/2022	30/04/2022
							F1203	Life Threatening Illness or Injury	A070504	Premature Discharge of Battery	01/04/2022	30/04/2022
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	01/04/2022	30/04/2022
							F23	Unexpected Medical Intervention	A070504	Premature Discharge of Battery	01/04/2022	30/04/2022
					CT981	Single use	E011903	Syncope/Fainting	A070504	Premature Discharge of Battery	01/04/2022	30/04/2022
							E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	01/04/2022	30/04/2022

							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	01/04/2022	30/04/2022
							F1203	Life Threatening Illness or Injury	A070504	Premature Discharge of Battery	01/04/2022	30/04/2022
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	01/04/2022	30/04/2022
							F23	Unexpected Medical Intervention	A070504	Premature Discharge of Battery	01/04/2022	30/04/2022
					CT983	Surgical invasive	E011903	Syncope/Fainting	A070504	Premature Discharge of Battery	01/04/2022	30/04/2022
							E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	01/04/2022	30/04/2022
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	01/04/2022	30/04/2022
							F1203	Life Threatening Illness or Injury	A070504	Premature Discharge of Battery	01/04/2022	30/04/2022
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	01/04/2022	30/04/2022
							F23	Unexpected Medical Intervention	A070504	Premature Discharge of Battery	01/04/2022	30/04/2022
6498	2021	All other reportable incidents	63188	Left-ventricular wireless pacing system sensor/transmitter	CT141	Electrodes/Leads and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A010102	Device Appears to Trigger Rejection		
					CT145	Cardiovascular devices	E2403	No Clinical Signs, Symptoms or Conditions	A010102	Device Appears to Trigger Rejection		
					CT1577	Wireless monitoring system transmitters	E2403	No Clinical Signs, Symptoms or Conditions	A010102	Device Appears to Trigger Rejection		
					CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A010102	Device Appears to Trigger Rejection		
					CT2317	Transmitters/Receivers/Transceivers	E2403	No Clinical Signs, Symptoms or Conditions	A010102	Device Appears to Trigger Rejection		
					CT233	Surgical	E2403	No Clinical Signs, Symptoms or Conditions	A010102	Device Appears to Trigger Rejection		
					CT2406	Implantable	E2403	No Clinical Signs, Symptoms or Conditions	A010102	Device Appears to Trigger Rejection		
					CT2407	Active implantable	E2403	No Clinical Signs, Symptoms or Conditions	A010102	Device Appears to Trigger Rejection		
					CT255	Transducers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A010102	Device Appears to Trigger Rejection		
					CT2726	Physiological monitoring/measuring devices	E2403	No Clinical Signs, Symptoms or Conditions	A010102	Device Appears to Trigger Rejection		
					CT2727	Active implant devices/ancillaries	E2403	No Clinical Signs, Symptoms or Conditions	A010102	Device Appears to Trigger Rejection		
					CT2750	Implantable cardiac electrical stimulators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A010102	Device Appears to Trigger Rejection		
					CT2885	Physiological monitoring/measuring system transmitters/receivers/transceivers	E2403	No Clinical Signs, Symptoms or Conditions	A010102	Device Appears to Trigger Rejection		
					CT2985	Active	E2403	No Clinical Signs, Symptoms or Conditions	A010102	Device Appears to Trigger Rejection		
					CT3001	Implantable metal	E2403	No Clinical Signs, Symptoms or Conditions	A010102	Device Appears to Trigger Rejection		
					CT3002	Implantable synthetic polymer	E2403	No Clinical Signs, Symptoms or Conditions	A010102	Device Appears to Trigger Rejection		
					CT3006	Implantable inorganic material	E2403	No Clinical Signs, Symptoms or Conditions	A010102	Device Appears to Trigger Rejection		
					CT314	Electrically powered	E2403	No Clinical Signs, Symptoms or Conditions	A010102	Device Appears to Trigger Rejection		
					CT321	Long-term surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A010102	Device Appears to Trigger Rejection		
					CT384	Sensors and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A010102	Device Appears to Trigger Rejection		
					CT449	Generators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A010102	Device Appears to Trigger Rejection		
					CT981	Single use	E2403	No Clinical Signs, Symptoms or Conditions	A010102	Device Appears to Trigger Rejection		
					CT983	Surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A010102	Device Appears to Trigger Rejection		
					CT994	Sensors	E2403	No Clinical Signs, Symptoms or Conditions	A010102	Device Appears to Trigger Rejection		
6499	2022	All other reportable incidents	60789	Intracardiac pacemaker	CT130	Catheters and associated devices	E060105	Ectopic Heartbeat	A051201	Device Dislodged or Dislocated	22/04/2022	22/04/2022
									A070101	Failure to Capture	22/04/2022	22/04/2022
									A070102	High Capture Threshold	22/04/2022	22/04/2022
							E060109	Tachycardia	A051201	Device Dislodged or Dislocated	22/04/2022	22/04/2022
									A070101	Failure to Capture	22/04/2022	22/04/2022
									A070102	High Capture Threshold	22/04/2022	22/04/2022
							E2403	No Clinical Signs, Symptoms or Conditions	A051201	Device Dislodged or Dislocated	22/04/2022	22/04/2022
									A070101	Failure to Capture	22/04/2022	22/04/2022
									A070102	High Capture Threshold	22/04/2022	22/04/2022
							F08	Hospitalization or Prolonged Hospitalization	A051201	Device Dislodged or Dislocated	22/04/2022	22/04/2022
									A070101	Failure to Capture	22/04/2022	22/04/2022
									A070102	High Capture Threshold	22/04/2022	22/04/2022
							F12	Serious Injury/ Illness/ Impairment	A051201	Device Dislodged or Dislocated	22/04/2022	22/04/2022
									A070101	Failure to Capture	22/04/2022	22/04/2022
									A070102	High Capture Threshold	22/04/2022	22/04/2022
							F1901	Additional Surgery	A051201	Device Dislodged or Dislocated	22/04/2022	22/04/2022
									A070101	Failure to Capture	22/04/2022	22/04/2022
									A070102	High Capture Threshold	22/04/2022	22/04/2022
							F1905	Device Revision or Replacement	A051201	Device Dislodged or Dislocated	22/04/2022	22/04/2022
									A070101	Failure to Capture	22/04/2022	22/04/2022
									A070102	High Capture Threshold	22/04/2022	22/04/2022
							F23	Unexpected Medical Intervention	A051201	Device Dislodged or Dislocated	22/04/2022	22/04/2022
									A070101	Failure to Capture	22/04/2022	22/04/2022
									A070102	High Capture Threshold	22/04/2022	22/04/2022
					CT145	Cardiovascular devices	E060105	Ectopic Heartbeat	A051201	Device Dislodged or Dislocated	22/04/2022	22/04/2022
									A070101	Failure to Capture	22/04/2022	22/04/2022
									A070102	High Capture Threshold	22/04/2022	22/04/2022
							E060109	Tachycardia	A051201	Device Dislodged or Dislocated	22/04/2022	22/04/2022
									A070101	Failure to Capture	22/04/2022	22/04/2022
									A070102	High Capture Threshold	22/04/2022	22/04/2022
							E2403	No Clinical Signs, Symptoms or Conditions	A051201	Device Dislodged or Dislocated	22/04/2022	22/04/2022
									A070101	Failure to Capture	22/04/2022	22/04/2022
									A070102	High Capture Threshold	22/04/2022	22/04/2022
							F08	Hospitalization or Prolonged Hospitalization	A051201	Device Dislodged or Dislocated	22/04/2022	22/04/2022
									A070101	Failure to Capture	22/04/2022	22/04/2022
									A070102	High Capture Threshold	22/04/2022	22/04/2022
							F12	Serious Injury/ Illness/ Impairment	A051201	Device Dislodged or Dislocated	22/04/2022	22/04/2022
									A070101	Failure to Capture	22/04/2022	22/04/2022
									A070102	High Capture Threshold	22/04/2022	22/04/2022
							F1901	Additional Surgery	A051201	Device Dislodged or Dislocated	22/04/2022	22/04/2022
									A070101	Failure to Capture	22/04/2022	22/04/2022
									A070102	High Capture Threshold	22/04/2022	22/04/2022
							F1905	Device Revision or Replacement	A051201	Device Dislodged or Dislocated	22/04/2022	22/04/2022
									A070101	Failure to Capture	22/04/2022	22/04/2022
									A070102	High Capture Threshold	22/04/2022	22/04/2022
							F23	Unexpected Medical Intervention	A051201	Device Dislodged or Dislocated	22/04/2022	22/04/2022
									A070101	Failure to Capture	22/04/2022	22/04/2022
									A070102	High Capture Threshold	22/04/2022	22/04/2022
					CT1684	Cardiac pacemakers and associated devices	E060105	Ectopic Heartbeat	A051201	Device Dislodged or Dislocated	22/04/2022	22/04/2022
									A070101	Failure to Capture	22/04/2022	22/04/2022
									A070102	High Capture Threshold	22/04/2022	22/04/2022
							E060109	Tachycardia	A051201	Device Dislodged or Dislocated	22/04/2022	22/04/2022

								A070101	Failure to Capture	22/04/2022	22/04/2022
								A070102	High Capture Threshold	22/04/2022	22/04/2022
							E2403	No Clinical Signs, Symptoms or Conditions	A051201	Device Dislodged or Dislocated	22/04/2022	22/04/2022
									A070101	Failure to Capture	22/04/2022	22/04/2022
									A070102	High Capture Threshold	22/04/2022	22/04/2022
							F08	Hospitalization or Prolonged Hospitalization	A051201	Device Dislodged or Dislocated	22/04/2022	22/04/2022
									A070101	Failure to Capture	22/04/2022	22/04/2022
									A070102	High Capture Threshold	22/04/2022	22/04/2022
							F12	Serious Injury/ Illness/ Impairment	A051201	Device Dislodged or Dislocated	22/04/2022	22/04/2022
									A070101	Failure to Capture	22/04/2022	22/04/2022
									A070102	High Capture Threshold	22/04/2022	22/04/2022
							F1901	Additional Surgery	A051201	Device Dislodged or Dislocated	22/04/2022	22/04/2022
									A070101	Failure to Capture	22/04/2022	22/04/2022
									A070102	High Capture Threshold	22/04/2022	22/04/2022
							F1905	Device Revision or Replacement	A051201	Device Dislodged or Dislocated	22/04/2022	22/04/2022
									A070101	Failure to Capture	22/04/2022	22/04/2022
									A070102	High Capture Threshold	22/04/2022	22/04/2022
							F23	Unexpected Medical Intervention	A051201	Device Dislodged or Dislocated	22/04/2022	22/04/2022
									A070101	Failure to Capture	22/04/2022	22/04/2022
									A070102	High Capture Threshold	22/04/2022	22/04/2022
						CT1685	Cardiac pacemakers	E060105	Ectopic Heartbeat	A051201	Device Dislodged or Dislocated	22/04/2022	22/04/2022
										A070101	Failure to Capture	22/04/2022	22/04/2022
										A070102	High Capture Threshold	22/04/2022	22/04/2022
								E060109	Tachycardia	A051201	Device Dislodged or Dislocated	22/04/2022	22/04/2022
										A070101	Failure to Capture	22/04/2022	22/04/2022
										A070102	High Capture Threshold	22/04/2022	22/04/2022
								E2403	No Clinical Signs, Symptoms or Conditions	A051201	Device Dislodged or Dislocated	22/04/2022	22/04/2022
										A070101	Failure to Capture	22/04/2022	22/04/2022
										A070102	High Capture Threshold	22/04/2022	22/04/2022
								F08	Hospitalization or Prolonged Hospitalization	A051201	Device Dislodged or Dislocated	22/04/2022	22/04/2022
										A070101	Failure to Capture	22/04/2022	22/04/2022
										A070102	High Capture Threshold	22/04/2022	22/04/2022
								F12	Serious Injury/ Illness/ Impairment	A051201	Device Dislodged or Dislocated	22/04/2022	22/04/2022
										A070101	Failure to Capture	22/04/2022	22/04/2022
										A070102	High Capture Threshold	22/04/2022	22/04/2022
								F1901	Additional Surgery	A051201	Device Dislodged or Dislocated	22/04/2022	22/04/2022
										A070101	Failure to Capture	22/04/2022	22/04/2022
										A070102	High Capture Threshold	22/04/2022	22/04/2022
								F1905	Device Revision or Replacement	A051201	Device Dislodged or Dislocated	22/04/2022	22/04/2022
										A070101	Failure to Capture	22/04/2022	22/04/2022
										A070102	High Capture Threshold	22/04/2022	22/04/2022
								F23	Unexpected Medical Intervention	A051201	Device Dislodged or Dislocated	22/04/2022	22/04/2022
										A070101	Failure to Capture	22/04/2022	22/04/2022
										A070102	High Capture Threshold	22/04/2022	22/04/2022
						CT233	Surgical	E060105	Ectopic Heartbeat	A051201	Device Dislodged or Dislocated	22/04/2022	22/04/2022
										A070101	Failure to Capture	22/04/2022	22/04/2022
										A070102	High Capture Threshold	22/04/2022	22/04/2022
								E060109	Tachycardia	A051201	Device Dislodged or Dislocated	22/04/2022	22/04/2022
										A070101	Failure to Capture	22/04/2022	22/04/2022
										A070102	High Capture Threshold	22/04/2022	22/04/2022
								E2403	No Clinical Signs, Symptoms or Conditions	A051201	Device Dislodged or Dislocated	22/04/2022	22/04/2022
										A070101	Failure to Capture	22/04/2022	22/04/2022
										A070102	High Capture Threshold	22/04/2022	22/04/2022
								F08	Hospitalization or Prolonged Hospitalization	A051201	Device Dislodged or Dislocated	22/04/2022	22/04/2022
										A070101	Failure to Capture	22/04/2022	22/04/2022
										A070102	High Capture Threshold	22/04/2022	22/04/2022
								F12	Serious Injury/ Illness/ Impairment	A051201	Device Dislodged or Dislocated	22/04/2022	22/04/2022
										A070101	Failure to Capture	22/04/2022	22/04/2022
										A070102	High Capture Threshold	22/04/2022	22/04/2022
								F1901	Additional Surgery	A051201	Device Dislodged or Dislocated	22/04/2022	22/04/2022
										A070101	Failure to Capture	22/04/2022	22/04/2022
										A070102	High Capture Threshold	22/04/2022	22/04/2022
								F1905	Device Revision or Replacement	A051201	Device Dislodged or Dislocated	22/04/2022	22/04/2022
										A070101	Failure to Capture	22/04/2022	22/04/2022
										A070102	High Capture Threshold	22/04/2022	22/04/2022
								F23	Unexpected Medical Intervention	A051201	Device Dislodged or Dislocated	22/04/2022	22/04/2022
										A070101	Failure to Capture	22/04/2022	22/04/2022
										A070102	High Capture Threshold	22/04/2022	22/04/2022
						CT2406	Implantable	E060105	Ectopic Heartbeat	A051201	Device Dislodged or Dislocated	22/04/2022	22/04/2022
										A070101	Failure to Capture	22/04/2022	22/04/2022
										A070102	High Capture Threshold	22/04/2022	22/04/2022
								E060109	Tachycardia	A051201	Device Dislodged or Dislocated	22/04/2022	22/04/2022
										A070101	Failure to Capture	22/04/2022	22/04/2022
										A070102	High Capture Threshold	22/04/2022	22/04/2022
								E2403	No Clinical Signs, Symptoms or Conditions	A051201	Device Dislodged or Dislocated	22/04/2022	22/04/2022
										A070101	Failure to Capture	22/04/2022	22/04/2022
										A070102	High Capture Threshold	22/04/2022	22/04/2022
								F08	Hospitalization or Prolonged Hospitalization	A051201	Device Dislodged or Dislocated	22/04/2022	22/04/2022
										A070101	Failure to Capture	22/04/2022	22/04/2022
										A070102	High Capture Threshold	22/04/2022	22/04/2022
								F12	Serious Injury/ Illness/ Impairment	A051201	Device Dislodged or Dislocated	22/04/2022	22/04/2022
										A070101	Failure to Capture	22/04/2022	22/04/2022
										A070102	High Capture Threshold	22/04/2022	22/04/2022
								F1901	Additional Surgery	A051201	Device Dislodged or Dislocated	22/04/2022	22/04/2022
										A070101	Failure to Capture	22/04/2022	22/04/2022
										A070102	High Capture Threshold	22/04/2022	22/04/2022
								F1905	Device Revision or Replacement	A051201	Device Dislodged or Dislocated	22/04/2022	22/04/2022
										A070101	Failure to Capture	22/04/2022	22/04/2022

								A070102	High Capture Threshold	22/04/2022	22/04/2022
						F23	Unexpected Medical Intervention	A051201	Device Dislodged or Dislocated	22/04/2022	22/04/2022
								A070101	Failure to Capture	22/04/2022	22/04/2022
								A070102	High Capture Threshold	22/04/2022	22/04/2022
					CT2407	Active implantable	E060105	Ectopic Heartbeat	A051201	Device Dislodged or Dislocated	22/04/2022	22/04/2022
									A070101	Failure to Capture	22/04/2022	22/04/2022
									A070102	High Capture Threshold	22/04/2022	22/04/2022
							E060109	Tachycardia	A051201	Device Dislodged or Dislocated	22/04/2022	22/04/2022
									A070101	Failure to Capture	22/04/2022	22/04/2022
									A070102	High Capture Threshold	22/04/2022	22/04/2022
							E2403	No Clinical Signs, Symptoms or Conditions	A051201	Device Dislodged or Dislocated	22/04/2022	22/04/2022
									A070101	Failure to Capture	22/04/2022	22/04/2022
									A070102	High Capture Threshold	22/04/2022	22/04/2022
							F08	Hospitalization or Prolonged Hospitalization	A051201	Device Dislodged or Dislocated	22/04/2022	22/04/2022
									A070101	Failure to Capture	22/04/2022	22/04/2022
									A070102	High Capture Threshold	22/04/2022	22/04/2022
							F12	Serious Injury/ Illness/ Impairment	A051201	Device Dislodged or Dislocated	22/04/2022	22/04/2022
									A070101	Failure to Capture	22/04/2022	22/04/2022
									A070102	High Capture Threshold	22/04/2022	22/04/2022
							F1901	Additional Surgery	A051201	Device Dislodged or Dislocated	22/04/2022	22/04/2022
									A070101	Failure to Capture	22/04/2022	22/04/2022
									A070102	High Capture Threshold	22/04/2022	22/04/2022
							F1905	Device Revision or Replacement	A051201	Device Dislodged or Dislocated	22/04/2022	22/04/2022
									A070101	Failure to Capture	22/04/2022	22/04/2022
									A070102	High Capture Threshold	22/04/2022	22/04/2022
							F23	Unexpected Medical Intervention	A051201	Device Dislodged or Dislocated	22/04/2022	22/04/2022
									A070101	Failure to Capture	22/04/2022	22/04/2022
									A070102	High Capture Threshold	22/04/2022	22/04/2022
					CT257	Implantable cardiac pacemakers	E060105	Ectopic Heartbeat	A051201	Device Dislodged or Dislocated	22/04/2022	22/04/2022
									A070101	Failure to Capture	22/04/2022	22/04/2022
									A070102	High Capture Threshold	22/04/2022	22/04/2022
							E060109	Tachycardia	A051201	Device Dislodged or Dislocated	22/04/2022	22/04/2022
									A070101	Failure to Capture	22/04/2022	22/04/2022
									A070102	High Capture Threshold	22/04/2022	22/04/2022
							E2403	No Clinical Signs, Symptoms or Conditions	A051201	Device Dislodged or Dislocated	22/04/2022	22/04/2022
									A070101	Failure to Capture	22/04/2022	22/04/2022
									A070102	High Capture Threshold	22/04/2022	22/04/2022
							F08	Hospitalization or Prolonged Hospitalization	A051201	Device Dislodged or Dislocated	22/04/2022	22/04/2022
									A070101	Failure to Capture	22/04/2022	22/04/2022
									A070102	High Capture Threshold	22/04/2022	22/04/2022
							F12	Serious Injury/ Illness/ Impairment	A051201	Device Dislodged or Dislocated	22/04/2022	22/04/2022
									A070101	Failure to Capture	22/04/2022	22/04/2022
									A070102	High Capture Threshold	22/04/2022	22/04/2022
							F1901	Additional Surgery	A051201	Device Dislodged or Dislocated	22/04/2022	22/04/2022
									A070101	Failure to Capture	22/04/2022	22/04/2022
									A070102	High Capture Threshold	22/04/2022	22/04/2022
							F1905	Device Revision or Replacement	A051201	Device Dislodged or Dislocated	22/04/2022	22/04/2022
									A070101	Failure to Capture	22/04/2022	22/04/2022
									A070102	High Capture Threshold	22/04/2022	22/04/2022
							F23	Unexpected Medical Intervention	A051201	Device Dislodged or Dislocated	22/04/2022	22/04/2022
									A070101	Failure to Capture	22/04/2022	22/04/2022
									A070102	High Capture Threshold	22/04/2022	22/04/2022
					CT2727	Active implant devices/ancillaries	E060105	Ectopic Heartbeat	A051201	Device Dislodged or Dislocated	22/04/2022	22/04/2022
									A070101	Failure to Capture	22/04/2022	22/04/2022
									A070102	High Capture Threshold	22/04/2022	22/04/2022
							E060109	Tachycardia	A051201	Device Dislodged or Dislocated	22/04/2022	22/04/2022
									A070101	Failure to Capture	22/04/2022	22/04/2022
									A070102	High Capture Threshold	22/04/2022	22/04/2022
							E2403	No Clinical Signs, Symptoms or Conditions	A051201	Device Dislodged or Dislocated	22/04/2022	22/04/2022
									A070101	Failure to Capture	22/04/2022	22/04/2022
									A070102	High Capture Threshold	22/04/2022	22/04/2022
							F08	Hospitalization or Prolonged Hospitalization	A051201	Device Dislodged or Dislocated	22/04/2022	22/04/2022
									A070101	Failure to Capture	22/04/2022	22/04/2022
									A070102	High Capture Threshold	22/04/2022	22/04/2022
							F12	Serious Injury/ Illness/ Impairment	A051201	Device Dislodged or Dislocated	22/04/2022	22/04/2022
									A070101	Failure to Capture	22/04/2022	22/04/2022
									A070102	High Capture Threshold	22/04/2022	22/04/2022
							F1901	Additional Surgery	A051201	Device Dislodged or Dislocated	22/04/2022	22/04/2022
									A070101	Failure to Capture	22/04/2022	22/04/2022
									A070102	High Capture Threshold	22/04/2022	22/04/2022
							F1905	Device Revision or Replacement	A051201	Device Dislodged or Dislocated	22/04/2022	22/04/2022
									A070101	Failure to Capture	22/04/2022	22/04/2022
									A070102	High Capture Threshold	22/04/2022	22/04/2022
							F23	Unexpected Medical Intervention	A051201	Device Dislodged or Dislocated	22/04/2022	22/04/2022
									A070101	Failure to Capture	22/04/2022	22/04/2022
									A070102	High Capture Threshold	22/04/2022	22/04/2022
					CT2750	Implantable cardiac electrical stimulators and associated devices	E060105	Ectopic Heartbeat	A051201	Device Dislodged or Dislocated	22/04/2022	22/04/2022
									A070101	Failure to Capture	22/04/2022	22/04/2022
									A070102	High Capture Threshold	22/04/2022	22/04/2022
							E060109	Tachycardia	A051201	Device Dislodged or Dislocated	22/04/2022	22/04/2022
									A070101	Failure to Capture	22/04/2022	22/04/2022
									A070102	High Capture Threshold	22/04/2022	22/04/2022
							E2403	No Clinical Signs, Symptoms or Conditions	A051201	Device Dislodged or Dislocated	22/04/2022	22/04/2022
									A070101	Failure to Capture	22/04/2022	22/04/2022
									A070102	High Capture Threshold	22/04/2022	22/04/2022
							F08	Hospitalization or Prolonged Hospitalization	A051201	Device Dislodged or Dislocated	22/04/2022	22/04/2022
									A070101	Failure to Capture	22/04/2022	22/04/2022
									A070102	High Capture Threshold	22/04/2022	22/04/2022

							F12	Serious Injury/ Illness/ Impairment	A051201	Device Dislodged or Dislocated	22/04/2022	22/04/2022
									A070101	Failure to Capture	22/04/2022	22/04/2022
									A070102	High Capture Threshold	22/04/2022	22/04/2022
							F1901	Additional Surgery	A051201	Device Dislodged or Dislocated	22/04/2022	22/04/2022
									A070101	Failure to Capture	22/04/2022	22/04/2022
									A070102	High Capture Threshold	22/04/2022	22/04/2022
							F1905	Device Revision or Replacement	A051201	Device Dislodged or Dislocated	22/04/2022	22/04/2022
									A070101	Failure to Capture	22/04/2022	22/04/2022
									A070102	High Capture Threshold	22/04/2022	22/04/2022
							F23	Unexpected Medical Intervention	A051201	Device Dislodged or Dislocated	22/04/2022	22/04/2022
									A070101	Failure to Capture	22/04/2022	22/04/2022
									A070102	High Capture Threshold	22/04/2022	22/04/2022
						CT2818	Catheterization/Cannulation/Fluid path devices	E060105	Ectopic Heartbeat	A051201	Device Dislodged or Dislocated	22/04/2022	22/04/2022
										A070101	Failure to Capture	22/04/2022	22/04/2022
										A070102	High Capture Threshold	22/04/2022	22/04/2022
								E060109	Tachycardia	A051201	Device Dislodged or Dislocated	22/04/2022	22/04/2022
										A070101	Failure to Capture	22/04/2022	22/04/2022
										A070102	High Capture Threshold	22/04/2022	22/04/2022
								E2403	No Clinical Signs, Symptoms or Conditions	A051201	Device Dislodged or Dislocated	22/04/2022	22/04/2022
										A070101	Failure to Capture	22/04/2022	22/04/2022
										A070102	High Capture Threshold	22/04/2022	22/04/2022
								F08	Hospitalization or Prolonged Hospitalization	A051201	Device Dislodged or Dislocated	22/04/2022	22/04/2022
										A070101	Failure to Capture	22/04/2022	22/04/2022
										A070102	High Capture Threshold	22/04/2022	22/04/2022
								F12	Serious Injury/ Illness/ Impairment	A051201	Device Dislodged or Dislocated	22/04/2022	22/04/2022
										A070101	Failure to Capture	22/04/2022	22/04/2022
										A070102	High Capture Threshold	22/04/2022	22/04/2022
								F1901	Additional Surgery	A051201	Device Dislodged or Dislocated	22/04/2022	22/04/2022
										A070101	Failure to Capture	22/04/2022	22/04/2022
										A070102	High Capture Threshold	22/04/2022	22/04/2022
								F1905	Device Revision or Replacement	A051201	Device Dislodged or Dislocated	22/04/2022	22/04/2022
										A070101	Failure to Capture	22/04/2022	22/04/2022
										A070102	High Capture Threshold	22/04/2022	22/04/2022
								F23	Unexpected Medical Intervention	A051201	Device Dislodged or Dislocated	22/04/2022	22/04/2022
										A070101	Failure to Capture	22/04/2022	22/04/2022
										A070102	High Capture Threshold	22/04/2022	22/04/2022
						CT2866	Implantable cardiac electrical stimulators	E060105	Ectopic Heartbeat	A051201	Device Dislodged or Dislocated	22/04/2022	22/04/2022
										A070101	Failure to Capture	22/04/2022	22/04/2022
										A070102	High Capture Threshold	22/04/2022	22/04/2022
								E060109	Tachycardia	A051201	Device Dislodged or Dislocated	22/04/2022	22/04/2022
										A070101	Failure to Capture	22/04/2022	22/04/2022
										A070102	High Capture Threshold	22/04/2022	22/04/2022
								E2403	No Clinical Signs, Symptoms or Conditions	A051201	Device Dislodged or Dislocated	22/04/2022	22/04/2022
										A070101	Failure to Capture	22/04/2022	22/04/2022
										A070102	High Capture Threshold	22/04/2022	22/04/2022
								F08	Hospitalization or Prolonged Hospitalization	A051201	Device Dislodged or Dislocated	22/04/2022	22/04/2022
										A070101	Failure to Capture	22/04/2022	22/04/2022
										A070102	High Capture Threshold	22/04/2022	22/04/2022
								F12	Serious Injury/ Illness/ Impairment	A051201	Device Dislodged or Dislocated	22/04/2022	22/04/2022
										A070101	Failure to Capture	22/04/2022	22/04/2022
										A070102	High Capture Threshold	22/04/2022	22/04/2022
								F1901	Additional Surgery	A051201	Device Dislodged or Dislocated	22/04/2022	22/04/2022
										A070101	Failure to Capture	22/04/2022	22/04/2022
										A070102	High Capture Threshold	22/04/2022	22/04/2022
								F1905	Device Revision or Replacement	A051201	Device Dislodged or Dislocated	22/04/2022	22/04/2022
										A070101	Failure to Capture	22/04/2022	22/04/2022
										A070102	High Capture Threshold	22/04/2022	22/04/2022
								F23	Unexpected Medical Intervention	A051201	Device Dislodged or Dislocated	22/04/2022	22/04/2022
										A070101	Failure to Capture	22/04/2022	22/04/2022
										A070102	High Capture Threshold	22/04/2022	22/04/2022
						CT2985	Active	E060105	Ectopic Heartbeat	A051201	Device Dislodged or Dislocated	22/04/2022	22/04/2022
										A070101	Failure to Capture	22/04/2022	22/04/2022
										A070102	High Capture Threshold	22/04/2022	22/04/2022
								E060109	Tachycardia	A051201	Device Dislodged or Dislocated	22/04/2022	22/04/2022
										A070101	Failure to Capture	22/04/2022	22/04/2022
										A070102	High Capture Threshold	22/04/2022	22/04/2022
								E2403	No Clinical Signs, Symptoms or Conditions	A051201	Device Dislodged or Dislocated	22/04/2022	22/04/2022
										A070101	Failure to Capture	22/04/2022	22/04/2022
										A070102	High Capture Threshold	22/04/2022	22/04/2022
								F08	Hospitalization or Prolonged Hospitalization	A051201	Device Dislodged or Dislocated	22/04/2022	22/04/2022
										A070101	Failure to Capture	22/04/2022	22/04/2022
										A070102	High Capture Threshold	22/04/2022	22/04/2022
								F12	Serious Injury/ Illness/ Impairment	A051201	Device Dislodged or Dislocated	22/04/2022	22/04/2022
										A070101	Failure to Capture	22/04/2022	22/04/2022
										A070102	High Capture Threshold	22/04/2022	22/04/2022
								F1901	Additional Surgery	A051201	Device Dislodged or Dislocated	22/04/2022	22/04/2022
										A070101	Failure to Capture	22/04/2022	22/04/2022
										A070102	High Capture Threshold	22/04/2022	22/04/2022
								F1905	Device Revision or Replacement	A051201	Device Dislodged or Dislocated	22/04/2022	22/04/2022
										A070101	Failure to Capture	22/04/2022	22/04/2022
										A070102	High Capture Threshold	22/04/2022	22/04/2022
								F23	Unexpected Medical Intervention	A051201	Device Dislodged or Dislocated	22/04/2022	22/04/2022
										A070101	Failure to Capture	22/04/2022	22/04/2022
										A070102	High Capture Threshold	22/04/2022	22/04/2022
						CT3001	Implantable metal	E060105	Ectopic Heartbeat	A051201	Device Dislodged or Dislocated	22/04/2022	22/04/2022
										A070101	Failure to Capture	22/04/2022	22/04/2022
										A070102	High Capture Threshold	22/04/2022	22/04/2022
								E060109	Tachycardia	A051201	Device Dislodged or Dislocated	22/04/2022	22/04/2022

								A070101	Failure to Capture	22/04/2022	22/04/2022
								A070102	High Capture Threshold	22/04/2022	22/04/2022
							E2403	No Clinical Signs, Symptoms or Conditions	A051201	Device Dislodged or Dislocated	22/04/2022	22/04/2022
									A070101	Failure to Capture	22/04/2022	22/04/2022
									A070102	High Capture Threshold	22/04/2022	22/04/2022
							F08	Hospitalization or Prolonged Hospitalization	A051201	Device Dislodged or Dislocated	22/04/2022	22/04/2022
									A070101	Failure to Capture	22/04/2022	22/04/2022
									A070102	High Capture Threshold	22/04/2022	22/04/2022
							F12	Serious Injury/ Illness/ Impairment	A051201	Device Dislodged or Dislocated	22/04/2022	22/04/2022
									A070101	Failure to Capture	22/04/2022	22/04/2022
									A070102	High Capture Threshold	22/04/2022	22/04/2022
							F1901	Additional Surgery	A051201	Device Dislodged or Dislocated	22/04/2022	22/04/2022
									A070101	Failure to Capture	22/04/2022	22/04/2022
									A070102	High Capture Threshold	22/04/2022	22/04/2022
							F1905	Device Revision or Replacement	A051201	Device Dislodged or Dislocated	22/04/2022	22/04/2022
									A070101	Failure to Capture	22/04/2022	22/04/2022
									A070102	High Capture Threshold	22/04/2022	22/04/2022
							F23	Unexpected Medical Intervention	A051201	Device Dislodged or Dislocated	22/04/2022	22/04/2022
									A070101	Failure to Capture	22/04/2022	22/04/2022
									A070102	High Capture Threshold	22/04/2022	22/04/2022
					CT3002	Implantable synthetic polymer	E060105	Ectopic Heartbeat	A051201	Device Dislodged or Dislocated	22/04/2022	22/04/2022
									A070101	Failure to Capture	22/04/2022	22/04/2022
									A070102	High Capture Threshold	22/04/2022	22/04/2022
							E060109	Tachycardia	A051201	Device Dislodged or Dislocated	22/04/2022	22/04/2022
									A070101	Failure to Capture	22/04/2022	22/04/2022
									A070102	High Capture Threshold	22/04/2022	22/04/2022
							E2403	No Clinical Signs, Symptoms or Conditions	A051201	Device Dislodged or Dislocated	22/04/2022	22/04/2022
									A070101	Failure to Capture	22/04/2022	22/04/2022
									A070102	High Capture Threshold	22/04/2022	22/04/2022
							F08	Hospitalization or Prolonged Hospitalization	A051201	Device Dislodged or Dislocated	22/04/2022	22/04/2022
									A070101	Failure to Capture	22/04/2022	22/04/2022
									A070102	High Capture Threshold	22/04/2022	22/04/2022
							F12	Serious Injury/ Illness/ Impairment	A051201	Device Dislodged or Dislocated	22/04/2022	22/04/2022
									A070101	Failure to Capture	22/04/2022	22/04/2022
									A070102	High Capture Threshold	22/04/2022	22/04/2022
							F1901	Additional Surgery	A051201	Device Dislodged or Dislocated	22/04/2022	22/04/2022
									A070101	Failure to Capture	22/04/2022	22/04/2022
									A070102	High Capture Threshold	22/04/2022	22/04/2022
							F1905	Device Revision or Replacement	A051201	Device Dislodged or Dislocated	22/04/2022	22/04/2022
									A070101	Failure to Capture	22/04/2022	22/04/2022
									A070102	High Capture Threshold	22/04/2022	22/04/2022
							F23	Unexpected Medical Intervention	A051201	Device Dislodged or Dislocated	22/04/2022	22/04/2022
									A070101	Failure to Capture	22/04/2022	22/04/2022
									A070102	High Capture Threshold	22/04/2022	22/04/2022
					CT3006	Implantable inorganic material	E060105	Ectopic Heartbeat	A051201	Device Dislodged or Dislocated	22/04/2022	22/04/2022
									A070101	Failure to Capture	22/04/2022	22/04/2022
									A070102	High Capture Threshold	22/04/2022	22/04/2022
							E060109	Tachycardia	A051201	Device Dislodged or Dislocated	22/04/2022	22/04/2022
									A070101	Failure to Capture	22/04/2022	22/04/2022
									A070102	High Capture Threshold	22/04/2022	22/04/2022
							E2403	No Clinical Signs, Symptoms or Conditions	A051201	Device Dislodged or Dislocated	22/04/2022	22/04/2022
									A070101	Failure to Capture	22/04/2022	22/04/2022
									A070102	High Capture Threshold	22/04/2022	22/04/2022
							F08	Hospitalization or Prolonged Hospitalization	A051201	Device Dislodged or Dislocated	22/04/2022	22/04/2022
									A070101	Failure to Capture	22/04/2022	22/04/2022
									A070102	High Capture Threshold	22/04/2022	22/04/2022
							F12	Serious Injury/ Illness/ Impairment	A051201	Device Dislodged or Dislocated	22/04/2022	22/04/2022
									A070101	Failure to Capture	22/04/2022	22/04/2022
									A070102	High Capture Threshold	22/04/2022	22/04/2022
							F1901	Additional Surgery	A051201	Device Dislodged or Dislocated	22/04/2022	22/04/2022
									A070101	Failure to Capture	22/04/2022	22/04/2022
									A070102	High Capture Threshold	22/04/2022	22/04/2022
							F1905	Device Revision or Replacement	A051201	Device Dislodged or Dislocated	22/04/2022	22/04/2022
									A070102	High Capture Threshold	22/04/2022	22/04/2022
							F23	Unexpected Medical Intervention	A051201	Device Dislodged or Dislocated	22/04/2022	22/04/2022
									A070101	Failure to Capture	22/04/2022	22/04/2022
									A070102	High Capture Threshold	22/04/2022	22/04/2022
					CT314	Electrically powered	E060105	Ectopic Heartbeat	A051201	Device Dislodged or Dislocated	22/04/2022	22/04/2022
									A070101	Failure to Capture	22/04/2022	22/04/2022
									A070102	High Capture Threshold	22/04/2022	22/04/2022
							E060109	Tachycardia	A051201	Device Dislodged or Dislocated	22/04/2022	22/04/2022
									A070101	Failure to Capture	22/04/2022	22/04/2022
									A070102	High Capture Threshold	22/04/2022	22/04/2022
							E2403	No Clinical Signs, Symptoms or Conditions	A051201	Device Dislodged or Dislocated	22/04/2022	22/04/2022
									A070101	Failure to Capture	22/04/2022	22/04/2022
									A070102	High Capture Threshold	22/04/2022	22/04/2022
							F08	Hospitalization or Prolonged Hospitalization	A051201	Device Dislodged or Dislocated	22/04/2022	22/04/2022
									A070101	Failure to Capture	22/04/2022	22/04/2022
									A070102	High Capture Threshold	22/04/2022	22/04/2022
							F12	Serious Injury/ Illness/ Impairment	A051201	Device Dislodged or Dislocated	22/04/2022	22/04/2022
									A070101	Failure to Capture	22/04/2022	22/04/2022
									A070102	High Capture Threshold	22/04/2022	22/04/2022
							F1901	Additional Surgery	A051201	Device Dislodged or Dislocated	22/04/2022	22/04/2022
									A070101	Failure to Capture	22/04/2022	22/04/2022
									A070102	High Capture Threshold	22/04/2022	22/04/2022
							F1905	Device Revision or Replacement	A051201	Device Dislodged or Dislocated	22/04/2022	22/04/2022
									A070101	Failure to Capture	22/04/2022	22/04/2022

									A070102	High Capture Threshold	22/04/2022	22/04/2022
							F23	Unexpected Medical Intervention	A051201	Device Dislodged or Dislocated	22/04/2022	22/04/2022
									A070101	Failure to Capture	22/04/2022	22/04/2022
									A070102	High Capture Threshold	22/04/2022	22/04/2022
					CT321	Long-term surgical invasive	E060105	Ectopic Heartbeat	A051201	Device Dislodged or Dislocated	22/04/2022	22/04/2022
									A070101	Failure to Capture	22/04/2022	22/04/2022
									A070102	High Capture Threshold	22/04/2022	22/04/2022
							E060109	Tachycardia	A051201	Device Dislodged or Dislocated	22/04/2022	22/04/2022
									A070101	Failure to Capture	22/04/2022	22/04/2022
									A070102	High Capture Threshold	22/04/2022	22/04/2022
							E2403	No Clinical Signs, Symptoms or Conditions	A051201	Device Dislodged or Dislocated	22/04/2022	22/04/2022
									A070101	Failure to Capture	22/04/2022	22/04/2022
									A070102	High Capture Threshold	22/04/2022	22/04/2022
							F08	Hospitalization or Prolonged Hospitalization	A051201	Device Dislodged or Dislocated	22/04/2022	22/04/2022
									A070101	Failure to Capture	22/04/2022	22/04/2022
									A070102	High Capture Threshold	22/04/2022	22/04/2022
							F12	Serious Injury/ Illness/ Impairment	A051201	Device Dislodged or Dislocated	22/04/2022	22/04/2022
									A070101	Failure to Capture	22/04/2022	22/04/2022
									A070102	High Capture Threshold	22/04/2022	22/04/2022
							F1901	Additional Surgery	A051201	Device Dislodged or Dislocated	22/04/2022	22/04/2022
									A070101	Failure to Capture	22/04/2022	22/04/2022
									A070102	High Capture Threshold	22/04/2022	22/04/2022
							F1905	Device Revision or Replacement	A051201	Device Dislodged or Dislocated	22/04/2022	22/04/2022
									A070101	Failure to Capture	22/04/2022	22/04/2022
									A070102	High Capture Threshold	22/04/2022	22/04/2022
							F23	Unexpected Medical Intervention	A051201	Device Dislodged or Dislocated	22/04/2022	22/04/2022
									A070101	Failure to Capture	22/04/2022	22/04/2022
									A070102	High Capture Threshold	22/04/2022	22/04/2022
					CT981	Single use	E060105	Ectopic Heartbeat	A051201	Device Dislodged or Dislocated	22/04/2022	22/04/2022
									A070101	Failure to Capture	22/04/2022	22/04/2022
									A070102	High Capture Threshold	22/04/2022	22/04/2022
							E060109	Tachycardia	A051201	Device Dislodged or Dislocated	22/04/2022	22/04/2022
									A070101	Failure to Capture	22/04/2022	22/04/2022
									A070102	High Capture Threshold	22/04/2022	22/04/2022
							E2403	No Clinical Signs, Symptoms or Conditions	A051201	Device Dislodged or Dislocated	22/04/2022	22/04/2022
									A070101	Failure to Capture	22/04/2022	22/04/2022
									A070102	High Capture Threshold	22/04/2022	22/04/2022
							F08	Hospitalization or Prolonged Hospitalization	A051201	Device Dislodged or Dislocated	22/04/2022	22/04/2022
									A070101	Failure to Capture	22/04/2022	22/04/2022
									A070102	High Capture Threshold	22/04/2022	22/04/2022
							F12	Serious Injury/ Illness/ Impairment	A051201	Device Dislodged or Dislocated	22/04/2022	22/04/2022
									A070101	Failure to Capture	22/04/2022	22/04/2022
									A070102	High Capture Threshold	22/04/2022	22/04/2022
							F1901	Additional Surgery	A051201	Device Dislodged or Dislocated	22/04/2022	22/04/2022
									A070101	Failure to Capture	22/04/2022	22/04/2022
									A070102	High Capture Threshold	22/04/2022	22/04/2022
							F1905	Device Revision or Replacement	A051201	Device Dislodged or Dislocated	22/04/2022	22/04/2022
									A070101	Failure to Capture	22/04/2022	22/04/2022
									A070102	High Capture Threshold	22/04/2022	22/04/2022
							F23	Unexpected Medical Intervention	A051201	Device Dislodged or Dislocated	22/04/2022	22/04/2022
									A070101	Failure to Capture	22/04/2022	22/04/2022
									A070102	High Capture Threshold	22/04/2022	22/04/2022
					CT983	Surgical invasive	E060105	Ectopic Heartbeat	A051201	Device Dislodged or Dislocated	22/04/2022	22/04/2022
									A070101	Failure to Capture	22/04/2022	22/04/2022
									A070102	High Capture Threshold	22/04/2022	22/04/2022
							E060109	Tachycardia	A051201	Device Dislodged or Dislocated	22/04/2022	22/04/2022
									A070101	Failure to Capture	22/04/2022	22/04/2022
									A070102	High Capture Threshold	22/04/2022	22/04/2022
							E2403	No Clinical Signs, Symptoms or Conditions	A051201	Device Dislodged or Dislocated	22/04/2022	22/04/2022
									A070101	Failure to Capture	22/04/2022	22/04/2022
									A070102	High Capture Threshold	22/04/2022	22/04/2022
							F08	Hospitalization or Prolonged Hospitalization	A051201	Device Dislodged or Dislocated	22/04/2022	22/04/2022
									A070101	Failure to Capture	22/04/2022	22/04/2022
									A070102	High Capture Threshold	22/04/2022	22/04/2022
							F12	Serious Injury/ Illness/ Impairment	A051201	Device Dislodged or Dislocated	22/04/2022	22/04/2022
									A070101	Failure to Capture	22/04/2022	22/04/2022
									A070102	High Capture Threshold	22/04/2022	22/04/2022
							F1901	Additional Surgery	A051201	Device Dislodged or Dislocated	22/04/2022	22/04/2022
									A070101	Failure to Capture	22/04/2022	22/04/2022
									A070102	High Capture Threshold	22/04/2022	22/04/2022
							F1905	Device Revision or Replacement	A051201	Device Dislodged or Dislocated	22/04/2022	22/04/2022
									A070101	Failure to Capture	22/04/2022	22/04/2022
									A070102	High Capture Threshold	22/04/2022	22/04/2022
							F23	Unexpected Medical Intervention	A051201	Device Dislodged or Dislocated	22/04/2022	22/04/2022
									A070101	Failure to Capture	22/04/2022	22/04/2022
									A070102	High Capture Threshold	22/04/2022	22/04/2022
6500	2021	All other reportable incidents	47270	Cardiac resynchronization therapy implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	03/09/2021	03/09/2021
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery	03/09/2021	03/09/2021
6501	2021	All other reportable incidents	35852	Single-chamber implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem	24/09/2021	24/09/2021
									A070504	Premature Discharge of Battery	24/09/2021	24/09/2021
							F08	Hospitalization or Prolonged Hospitalization	A0705	Battery Problem	24/09/2021	24/09/2021
									A070504	Premature Discharge of Battery	24/09/2021	24/09/2021
							F1905	Device Revision or Replacement	A0705	Battery Problem	24/09/2021	24/09/2021
									A070504	Premature Discharge of Battery	24/09/2021	24/09/2021
6502	2021	All other reportable incidents	35852	Single-chamber implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	06/09/2021	06/09/2021
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	06/09/2021	06/09/2021
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	06/09/2021	06/09/2021
6503	2021	All other reportable incidents	35852	Single-chamber implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem	29/09/2021	29/09/2021

									A070504	Premature Discharge of Battery	29/09/2021	29/09/2021
						F08	Hospitalization or Prolonged Hospitalization	A0705	Battery Problem	29/09/2021	29/09/2021
								A070504	Premature Discharge of Battery	29/09/2021	29/09/2021
						F1905	Device Revision or Replacement	A0705	Battery Problem	29/09/2021	29/09/2021
								A070504	Premature Discharge of Battery	29/09/2021	29/09/2021
6504	2021	All other reportable incidents	35852	Single-chamber implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem	27/08/2021	27/08/2021
									A070504	Premature Discharge of Battery	27/08/2021	27/08/2021
						F08	Hospitalization or Prolonged Hospitalization	A0705	Battery Problem	27/08/2021	27/08/2021
								A070504	Premature Discharge of Battery	27/08/2021	27/08/2021
						F1905	Device Revision or Replacement	A0705	Battery Problem	27/08/2021	27/08/2021
								A070504	Premature Discharge of Battery	27/08/2021	27/08/2021
6505	2021	All other reportable incidents	35852	Single-chamber implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem	06/10/2021	06/10/2021
									A070504	Premature Discharge of Battery	06/10/2021	06/10/2021
						F08	Hospitalization or Prolonged Hospitalization	A0705	Battery Problem	06/10/2021	06/10/2021
								A070504	Premature Discharge of Battery	06/10/2021	06/10/2021
						F1905	Device Revision or Replacement	A0705	Battery Problem	06/10/2021	06/10/2021
								A070504	Premature Discharge of Battery	06/10/2021	06/10/2021
6506	2021	All other reportable incidents	47265	Dual-chamber implantable pacemaker, rate-responsive	CT141	Electrodes/Leads and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem		
									A070910	Undersensing		
							F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A0705	Battery Problem		
									A070910	Undersensing		
					CT145	Cardiovascular devices	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem		
									A070910	Undersensing		
							F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A0705	Battery Problem		
									A070910	Undersensing		
					CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem		
									A070910	Undersensing		
							F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A0705	Battery Problem		
									A070910	Undersensing		
					CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem		
									A070910	Undersensing		
							F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A0705	Battery Problem		
									A070910	Undersensing		
					CT233	Surgical	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem		
									A070910	Undersensing		
							F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A0705	Battery Problem		
									A070910	Undersensing		
					CT2406	Implantable	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem		
									A070910	Undersensing		
							F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A0705	Battery Problem		
									A070910	Undersensing		
					CT2407	Active implantable	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem		
									A070910	Undersensing		
							F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A0705	Battery Problem		
									A070910	Undersensing		
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem		
									A070910	Undersensing		
							F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A0705	Battery Problem		
									A070910	Undersensing		
					CT2727	Active implant devices/ancillaries	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem		
									A070910	Undersensing		
							F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A0705	Battery Problem		
									A070910	Undersensing		
					CT2750	Implantable cardiac electrical stimulators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem		
									A070910	Undersensing		
							F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A0705	Battery Problem		
									A070910	Undersensing		
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem		
									A070910	Undersensing		
							F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A0705	Battery Problem		
									A070910	Undersensing		
					CT2985	Active	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem		
									A070910	Undersensing		
							F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A0705	Battery Problem		
									A070910	Undersensing		
					CT3001	Implantable metal	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem		
									A070910	Undersensing		
							F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A0705	Battery Problem		
									A070910	Undersensing		
					CT3002	Implantable synthetic polymer	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem		
									A070910	Undersensing		
							F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A0705	Battery Problem		
									A070910	Undersensing		
					CT3006	Implantable inorganic material	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem		
									A070910	Undersensing		
							F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A0705	Battery Problem		
									A070910	Undersensing		
					CT314	Electrically powered	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem		
									A070910	Undersensing		
							F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A0705	Battery Problem		
									A070910	Undersensing		
					CT321	Long-term surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem		
									A070910	Undersensing		
							F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A0705	Battery Problem		
									A070910	Undersensing		
					CT449	Generators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem		
									A070910	Undersensing		
							F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A0705	Battery Problem		

									A070910	Undersensing		
					CT981	Single use	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem		
									A070910	Undersensing		
							F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A0705	Battery Problem		
									A070910	Undersensing		
					CT983	Surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem		
									A070910	Undersensing		
							F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A0705	Battery Problem		
									A070910	Undersensing		
6507	2021	All other reportable incidents	35852	Single-chamber implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem	28/09/2021	28/09/2021
									A070504	Premature Discharge of Battery	28/09/2021	28/09/2021
							F08	Hospitalization or Prolonged Hospitalization	A0705	Battery Problem	28/09/2021	28/09/2021
									A070504	Premature Discharge of Battery	28/09/2021	28/09/2021
							F1905	Device Revision or Replacement	A0705	Battery Problem	28/09/2021	28/09/2021
									A070504	Premature Discharge of Battery	28/09/2021	28/09/2021
6508	2021	All other reportable incidents	35852	Single-chamber implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem	19/11/2021	19/11/2021
									A070504	Premature Discharge of Battery	19/11/2021	19/11/2021
							F08	Hospitalization or Prolonged Hospitalization	A0705	Battery Problem	19/11/2021	19/11/2021
									A070504	Premature Discharge of Battery	19/11/2021	19/11/2021
							F1903	Device Explantation	A0705	Battery Problem	19/11/2021	19/11/2021
									A070504	Premature Discharge of Battery	19/11/2021	19/11/2021
6509	2021	All other reportable incidents	36241	Pacemaker/defibrillator lead	CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A05	Mechanical Problem	27/08/2021	27/08/2021
									A070101	Failure to Capture	27/08/2021	27/08/2021
									A072201	High Impedance	27/08/2021	27/08/2021
							F08	Hospitalization or Prolonged Hospitalization	A05	Mechanical Problem	27/08/2021	27/08/2021
									A070101	Failure to Capture	27/08/2021	27/08/2021
									A072201	High Impedance	27/08/2021	27/08/2021
							F1905	Device Revision or Replacement	A05	Mechanical Problem	27/08/2021	27/08/2021
									A070101	Failure to Capture	27/08/2021	27/08/2021
									A072201	High Impedance	27/08/2021	27/08/2021
6510	2021	All other reportable incidents	35852	Single-chamber implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem	28/10/2021	28/10/2021
									A070504	Premature Discharge of Battery	28/10/2021	28/10/2021
							F08	Hospitalization or Prolonged Hospitalization	A0705	Battery Problem	28/10/2021	28/10/2021
									A070504	Premature Discharge of Battery	28/10/2021	28/10/2021
							F1905	Device Revision or Replacement	A0705	Battery Problem	28/10/2021	28/10/2021
									A070504	Premature Discharge of Battery	28/10/2021	28/10/2021
6511	2021	All other reportable incidents	37265	Dual-chamber implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	28/09/2021	28/09/2021
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	28/09/2021	28/09/2021
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	28/09/2021	28/09/2021
6512	2022	All other reportable incidents	45966	Transthoracic pacing lead	CT1684	Cardiac pacemakers and associated devices	E060104	Bradycardia	A0401	Break		
							E2403	No Clinical Signs, Symptoms or Conditions	A0401	Break		
							F19	Surgical Intervention	A0401	Break		
6513	2022	All other reportable incidents	45966	Transthoracic pacing lead	CT1684	Cardiac pacemakers and associated devices	E060104	Bradycardia	A0401	Break		
							E2403	No Clinical Signs, Symptoms or Conditions	A0401	Break		
							F19	Surgical Intervention	A0401	Break		
6514	2022	All other reportable incidents	45966	Transthoracic pacing lead	CT1684	Cardiac pacemakers and associated devices	E060104	Bradycardia	A0401	Break		
							E0602	Cardiac Arrest	A0401	Break		
							E2403	No Clinical Signs, Symptoms or Conditions	A0401	Break		
							F19	Surgical Intervention	A0401	Break		
							F2306	Resuscitation	A0401	Break		
6515	2022	All other reportable incidents	36241	Pacemaker/defibrillator lead	CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A040101	Fracture		
									A070911	Sensing Intermittently		
							F1001	Absence of Treatment	A040101	Fracture		
									A070911	Sensing Intermittently		
							F23	Unexpected Medical Intervention	A040101	Fracture		
									A070911	Sensing Intermittently		
6516	2022	All other reportable incidents	36241	Pacemaker/defibrillator lead	CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A072201	High Impedance		
							F26	No Health Consequences or Impact	A072201	High Impedance		
6517	2022	All other reportable incidents	17882	Physiologic-monitoring defibrillation system	CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A1603	Failure of Device to Self-Test		
							F27	Problem Identified During Non-Clinical Procedure	A1603	Failure of Device to Self-Test		
6518	2021	All other reportable incidents	36241	Pacemaker/defibrillator lead	CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A051201	Device Dislodged or Dislocated		
									A070101	Failure to Capture		
									A070102	High Capture Threshold		
							F26	No Health Consequences or Impact	A051201	Device Dislodged or Dislocated		
									A070101	Failure to Capture		
									A070102	High Capture Threshold		
6519	2022	All other reportable incidents	36241	Pacemaker/defibrillator lead	CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A072201	High Impedance		
							F26	No Health Consequences or Impact	A072201	High Impedance		
6520	2021	All other reportable incidents	17882	Physiologic-monitoring defibrillation system	CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A090811	Unable to Obtain Readings		
							F27	Problem Identified During Non-Clinical Procedure	A090811	Unable to Obtain Readings		
6521	2021	Death	17882	Physiologic-monitoring defibrillation system	CT1684	Cardiac pacemakers and associated devices	E0518	Unspecified Vascular Problem	A090801	Signal Artifact/Noise		
							F02	Death	A090801	Signal Artifact/Noise		
6522	2022	All other reportable incidents	17882	Physiologic-monitoring defibrillation system	CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A0201	Product Quality Problem		
									A2302	Use of Incorrect Control/Treatment Settings		
6523	2022	All other reportable incidents	36241	Pacemaker/defibrillator lead	CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A072201	High Impedance	04/08/2022	04/08/2022
							F08	Hospitalization or Prolonged Hospitalization	A072201	High Impedance	04/08/2022	04/08/2022
							F1905	Device Revision or Replacement	A072201	High Impedance	04/08/2022	04/08/2022
6524	2022	All other reportable incidents	35852	Single-chamber implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem	04/05/2022	04/05/2022
									A070504	Premature Discharge of Battery	04/05/2022	04/05/2022
							F08	Hospitalization or Prolonged Hospitalization	A0705	Battery Problem	04/05/2022	04/05/2022
									A070504	Premature Discharge of Battery	04/05/2022	04/05/2022
							F12	Serious Injury/ Illness/ Impairment	A0705	Battery Problem	04/05/2022	04/05/2022
									A070504	Premature Discharge of Battery	04/05/2022	04/05/2022
							F1905	Device Revision or Replacement	A0705	Battery Problem	04/05/2022	04/05/2022
									A070504	Premature Discharge of Battery	04/05/2022	04/05/2022
6525	2022	All other reportable incidents	35852	Single-chamber implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A07	Electrical /Electronic Property Problem	07/05/2022	07/05/2022
									A070504	Premature Discharge of Battery	07/05/2022	07/05/2022
									A070602	Delayed Charge Time	07/05/2022	07/05/2022

							F26	No Health Consequences or Impact	A07	Electrical /Electronic Property Problem	07/05/2022	07/05/2022
									A070504	Premature Discharge of Battery	07/05/2022	07/05/2022
									A070602	Delayed Charge Time	07/05/2022	07/05/2022
6526	2021	All other reportable incidents	36241	Pacemaker/defibrillator lead	CT1684	Cardiac pacemakers and associated devices	E2104	Electric Shock	A04	Material Integrity Problem		
									A070909	Oversensing		
									A071302	Inappropriate/Inadequate Shock/Stimulation		
									A090801	Signal Artifact/Noise		
									A1301	Failure to Read Input Signal		
							E2403	No Clinical Signs, Symptoms or Conditions	A04	Material Integrity Problem		
									A070909	Oversensing		
									A071302	Inappropriate/Inadequate Shock/Stimulation		
									A090801	Signal Artifact/Noise		
									A1301	Failure to Read Input Signal		
							F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A04	Material Integrity Problem		
									A070909	Oversensing		
									A071302	Inappropriate/Inadequate Shock/Stimulation		
									A090801	Signal Artifact/Noise		
									A1301	Failure to Read Input Signal		
6527	2022	All other reportable incidents	47265	Dual-chamber implantable pacemaker, rate-responsive	CT141	Electrodes/Leads and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem		
							F26	No Health Consequences or Impact	A0705	Battery Problem		
					CT145	Cardiovascular devices	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem		
							F26	No Health Consequences or Impact	A0705	Battery Problem		
					CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem		
							F26	No Health Consequences or Impact	A0705	Battery Problem		
					CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem		
							F26	No Health Consequences or Impact	A0705	Battery Problem		
					CT233	Surgical	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem		
							F26	No Health Consequences or Impact	A0705	Battery Problem		
					CT2406	Implantable	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem		
							F26	No Health Consequences or Impact	A0705	Battery Problem		
					CT2407	Active implantable	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem		
							F26	No Health Consequences or Impact	A0705	Battery Problem		
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem		
							F26	No Health Consequences or Impact	A0705	Battery Problem		
					CT2727	Active implant devices/ancillaries	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem		
							F26	No Health Consequences or Impact	A0705	Battery Problem		
					CT2750	Implantable cardiac electrical stimulators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem		
							F26	No Health Consequences or Impact	A0705	Battery Problem		
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem		
							F26	No Health Consequences or Impact	A0705	Battery Problem		
					CT2985	Active	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem		
							F26	No Health Consequences or Impact	A0705	Battery Problem		
					CT3001	Implantable metal	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem		
							F26	No Health Consequences or Impact	A0705	Battery Problem		
					CT3002	Implantable synthetic polymer	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem		
							F26	No Health Consequences or Impact	A0705	Battery Problem		
					CT3006	Implantable inorganic material	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem		
							F26	No Health Consequences or Impact	A0705	Battery Problem		
					CT314	Electrically powered	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem		
							F26	No Health Consequences or Impact	A0705	Battery Problem		
					CT321	Long-term surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem		
							F26	No Health Consequences or Impact	A0705	Battery Problem		
					CT449	Generators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem		
							F26	No Health Consequences or Impact	A0705	Battery Problem		
					CT981	Single use	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem		
							F26	No Health Consequences or Impact	A0705	Battery Problem		
					CT983	Surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem		
							F26	No Health Consequences or Impact	A0705	Battery Problem		
6528	2022	All other reportable incidents	35852	Single-chamber implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E2104	Electric Shock	A071302	Inappropriate/Inadequate Shock/Stimulation		
							E2403	No Clinical Signs, Symptoms or Conditions	A071302	Inappropriate/Inadequate Shock/Stimulation		
							F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A071302	Inappropriate/Inadequate Shock/Stimulation		
6529	2021	All other reportable incidents	47270	Cardiac resynchronization therapy implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A070101	Failure to Capture		
									A070909	Oversensing		
							F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A070101	Failure to Capture		
									A070909	Oversensing		
							F23	Unexpected Medical Intervention	A070101	Failure to Capture		
									A070909	Oversensing		
6530	2021	All other reportable incidents	37265	Dual-chamber implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E2104	Electric Shock	A0712	Pacing Problem		
									A071302	Inappropriate/Inadequate Shock/Stimulation		
							E2403	No Clinical Signs, Symptoms or Conditions	A0712	Pacing Problem		
									A071302	Inappropriate/Inadequate Shock/Stimulation		
							F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A0712	Pacing Problem		
									A071302	Inappropriate/Inadequate Shock/Stimulation		
6531	2021	All other reportable incidents	47265	Dual-chamber implantable pacemaker, rate-responsive	CT141	Electrodes/Leads and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F27	Problem Identified During Non-Clinical Procedure	A070504	Premature Discharge of Battery		
					CT145	Cardiovascular devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F27	Problem Identified During Non-Clinical Procedure	A070504	Premature Discharge of Battery		
					CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F27	Problem Identified During Non-Clinical Procedure	A070504	Premature Discharge of Battery		
					CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F27	Problem Identified During Non-Clinical Procedure	A070504	Premature Discharge of Battery		
					CT233	Surgical	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F27	Problem Identified During Non-Clinical Procedure	A070504	Premature Discharge of Battery		
					CT2406	Implantable	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F27	Problem Identified During Non-Clinical Procedure	A070504	Premature Discharge of Battery		
					CT2407	Active implantable	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F27	Problem Identified During Non-Clinical Procedure	A070504	Premature Discharge of Battery		
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		

						F27	Problem Identified During Non-Clinical Procedure	A070504	Premature Discharge of Battery		
				CT2727	Active implant devices/ancillaries	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
						F27	Problem Identified During Non-Clinical Procedure	A070504	Premature Discharge of Battery		
				CT2750	Implantable cardiac electrical stimulators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
						F27	Problem Identified During Non-Clinical Procedure	A070504	Premature Discharge of Battery		
				CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
						F27	Problem Identified During Non-Clinical Procedure	A070504	Premature Discharge of Battery		
				CT2985	Active	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
						F27	Problem Identified During Non-Clinical Procedure	A070504	Premature Discharge of Battery		
				CT3001	Implantable metal	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
						F27	Problem Identified During Non-Clinical Procedure	A070504	Premature Discharge of Battery		
				CT3002	Implantable synthetic polymer	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
						F27	Problem Identified During Non-Clinical Procedure	A070504	Premature Discharge of Battery		
				CT3006	Implantable inorganic material	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
						F27	Problem Identified During Non-Clinical Procedure	A070504	Premature Discharge of Battery		
				CT314	Electrically powered	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
						F27	Problem Identified During Non-Clinical Procedure	A070504	Premature Discharge of Battery		
				CT321	Long-term surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
						F27	Problem Identified During Non-Clinical Procedure	A070504	Premature Discharge of Battery		
				CT449	Generators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
						F27	Problem Identified During Non-Clinical Procedure	A070504	Premature Discharge of Battery		
				CT981	Single use	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
						F27	Problem Identified During Non-Clinical Procedure	A070504	Premature Discharge of Battery		
				CT983	Surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
						F27	Problem Identified During Non-Clinical Procedure	A070504	Premature Discharge of Battery		
6532	2022	All other reportable incidents	47270	Cardiac resynchronization therapy implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A071302	Inappropriate/Inadequate Shock/Stimulation		
									A23	Use of Device Problem		
							F23	Unexpected Medical Intervention	A071302	Inappropriate/Inadequate Shock/Stimulation		
									A23	Use of Device Problem		
6533	2022	All other reportable incidents	17882	Physiologic-monitoring defibrillation system	CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A1603	Failure of Device to Self-Test		
							F27	Problem Identified During Non-Clinical Procedure	A1603	Failure of Device to Self-Test		
6534	2022	All other reportable incidents	47265	Dual-chamber implantable pacemaker, rate-responsive	CT141	Electrodes/Leads and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
				CT145	Cardiovascular devices	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
				CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
				CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
				CT233	Surgical	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
				CT2406	Implantable	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
				CT2407	Active implantable	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
				CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
				CT2727	Active implant devices/ancillaries	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
				CT2750	Implantable cardiac electrical stimulators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
				CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
				CT2985	Active	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
				CT3001	Implantable metal	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
				CT3002	Implantable synthetic polymer	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
				CT3006	Implantable inorganic material	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem		

									A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
					CT314	Electrically powered	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
					CT321	Long-term surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
					CT449	Generators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
					CT981	Single use	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
					CT983	Surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
6535	2022	All other reportable incidents	36241	Pacemaker/defibrillator lead	CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A0203	Defective Device		
									A070101	Failure to Capture		
									A070102	High Capture Threshold		
									A072202	Low Impedance		
							F08	Hospitalization or Prolonged Hospitalization	A0203	Defective Device		
									A070101	Failure to Capture		
									A070102	High Capture Threshold		
									A072202	Low Impedance		
							F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A0203	Defective Device		
									A070101	Failure to Capture		
									A070102	High Capture Threshold		
									A072202	Low Impedance		
							F1905	Device Revision or Replacement	A0203	Defective Device		
									A070101	Failure to Capture		
									A070102	High Capture Threshold		
									A072202	Low Impedance		
6536	2022	All other reportable incidents	17882	Physiologic-monitoring defibrillation system	CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A071301	Failure to Deliver Shock/Stimulation		
							F26	No Health Consequences or Impact	A071301	Failure to Deliver Shock/Stimulation		
6537	2021	All other reportable incidents	36241	Pacemaker/defibrillator lead	CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A051201	Device Dislodged or Dislocated	25/11/2021	25/11/2021
									A070101	Failure to Capture	25/11/2021	25/11/2021
									A070909	Oversensing	25/11/2021	25/11/2021
							F08	Hospitalization or Prolonged Hospitalization	A051201	Device Dislodged or Dislocated	25/11/2021	25/11/2021
									A070101	Failure to Capture	25/11/2021	25/11/2021
									A070909	Oversensing	25/11/2021	25/11/2021
							F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A051201	Device Dislodged or Dislocated	25/11/2021	25/11/2021
									A070101	Failure to Capture	25/11/2021	25/11/2021
									A070909	Oversensing	25/11/2021	25/11/2021
							F15	Recognized Device or Procedural Complication	A051201	Device Dislodged or Dislocated	25/11/2021	25/11/2021
									A070101	Failure to Capture	25/11/2021	25/11/2021
									A070909	Oversensing	25/11/2021	25/11/2021
							F1905	Device Revision or Replacement	A051201	Device Dislodged or Dislocated	25/11/2021	25/11/2021
									A070101	Failure to Capture	25/11/2021	25/11/2021
									A070909	Oversensing	25/11/2021	25/11/2021
6538	2021	All other reportable incidents	37265	Dual-chamber implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A070909	Oversensing		
							F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A070909	Oversensing		
6539	2021	All other reportable incidents	37265	Dual-chamber implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
6540	2021	All other reportable incidents	47265	Dual-chamber implantable pacemaker, rate-responsive	CT141	Electrodes/Leads and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
					CT145	Cardiovascular devices	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
					CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
					CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
					CT233	Surgical	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
					CT2406	Implantable	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
					CT2407	Active implantable	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		

					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
					CT2727	Active implant devices/ancillaries	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
					CT2750	Implantable cardiac electrical stimulators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
					CT2985	Active	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
					CT3001	Implantable metal	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
					CT3002	Implantable synthetic polymer	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
					CT3006	Implantable inorganic material	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
					CT314	Electrically powered	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
					CT321	Long-term surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
					CT449	Generators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
					CT981	Single use	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
					CT983	Surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
6541	2022	All other reportable incidents	36241	Pacemaker/defibrillator lead	CT1684	Cardiac pacemakers and associated devices	E2104	Electric Shock	A070909	Oversensing		
									A071302	Inappropriate/Inadequate Shock/Stimulation		
									A090801	Signal Artifact/Noise		
									A1301	Failure to Read Input Signal		
							E2403	No Clinical Signs, Symptoms or Conditions	A070909	Oversensing		
									A071302	Inappropriate/Inadequate Shock/Stimulation		
									A090801	Signal Artifact/Noise		
									A1301	Failure to Read Input Signal		
							F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A070909	Oversensing		
									A071302	Inappropriate/Inadequate Shock/Stimulation		
									A090801	Signal Artifact/Noise		
									A1301	Failure to Read Input Signal		
6542	2022	All other reportable incidents	37265	Dual-chamber implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E060110	Ventricular Fibrillation	A070910	Undersensing		
									A071202	Inaccurate Synchronization		
							E2403	No Clinical Signs, Symptoms or Conditions	A070910	Undersensing		
									A071202	Inaccurate Synchronization		
							F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A070910	Undersensing		
									A071202	Inaccurate Synchronization		
							F15	Recognized Device or Procedural Complication	A070910	Undersensing		
									A071202	Inaccurate Synchronization		
6543	2022	All other reportable incidents	47265	Dual-chamber implantable pacemaker, rate-responsive	CT141	Electrodes/Leads and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	07/05/2022	07/05/2022
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	07/05/2022	07/05/2022
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	07/05/2022	07/05/2022
					CT145	Cardiovascular devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	07/05/2022	07/05/2022
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	07/05/2022	07/05/2022
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	07/05/2022	07/05/2022
					CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	07/05/2022	07/05/2022
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	07/05/2022	07/05/2022
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	07/05/2022	07/05/2022
					CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	07/05/2022	07/05/2022
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	07/05/2022	07/05/2022
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	07/05/2022	07/05/2022
					CT233	Surgical	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	07/05/2022	07/05/2022
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	07/05/2022	07/05/2022
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	07/05/2022	07/05/2022
					CT2406	Implantable	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	07/05/2022	07/05/2022

							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	07/05/2022	07/05/2022
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	07/05/2022	07/05/2022
					CT2407	Active implantable	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	07/05/2022	07/05/2022
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	07/05/2022	07/05/2022
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	07/05/2022	07/05/2022
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	07/05/2022	07/05/2022
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	07/05/2022	07/05/2022
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	07/05/2022	07/05/2022
					CT2727	Active implant devices/ancillaries	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	07/05/2022	07/05/2022
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	07/05/2022	07/05/2022
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	07/05/2022	07/05/2022
					CT2750	Implantable cardiac electrical stimulators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	07/05/2022	07/05/2022
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	07/05/2022	07/05/2022
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	07/05/2022	07/05/2022
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	07/05/2022	07/05/2022
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	07/05/2022	07/05/2022
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	07/05/2022	07/05/2022
					CT2985	Active	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	07/05/2022	07/05/2022
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	07/05/2022	07/05/2022
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	07/05/2022	07/05/2022
					CT3001	Implantable metal	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	07/05/2022	07/05/2022
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	07/05/2022	07/05/2022
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	07/05/2022	07/05/2022
					CT3002	Implantable synthetic polymer	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	07/05/2022	07/05/2022
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	07/05/2022	07/05/2022
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	07/05/2022	07/05/2022
					CT3006	Implantable inorganic material	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	07/05/2022	07/05/2022
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	07/05/2022	07/05/2022
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	07/05/2022	07/05/2022
					CT314	Electrically powered	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	07/05/2022	07/05/2022
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	07/05/2022	07/05/2022
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	07/05/2022	07/05/2022
					CT321	Long-term surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	07/05/2022	07/05/2022
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	07/05/2022	07/05/2022
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	07/05/2022	07/05/2022
					CT449	Generators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	07/05/2022	07/05/2022
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	07/05/2022	07/05/2022
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	07/05/2022	07/05/2022
					CT981	Single use	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	07/05/2022	07/05/2022
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	07/05/2022	07/05/2022
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	07/05/2022	07/05/2022
					CT983	Surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	07/05/2022	07/05/2022
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	07/05/2022	07/05/2022
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	07/05/2022	07/05/2022
6544	2022	All other reportable incidents	47265	Dual-chamber implantable pacemaker, rate-responsive	CT141	Electrodes/Leads and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem	28/06/2022	28/06/2022
									A070504	Premature Discharge of Battery	28/06/2022	28/06/2022
							F08	Hospitalization or Prolonged Hospitalization	A0705	Battery Problem	28/06/2022	28/06/2022
									A070504	Premature Discharge of Battery	28/06/2022	28/06/2022
							F1905	Device Revision or Replacement	A0705	Battery Problem	28/06/2022	28/06/2022
									A070504	Premature Discharge of Battery	28/06/2022	28/06/2022
					CT145	Cardiovascular devices	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem	28/06/2022	28/06/2022
									A070504	Premature Discharge of Battery	28/06/2022	28/06/2022
							F08	Hospitalization or Prolonged Hospitalization	A0705	Battery Problem	28/06/2022	28/06/2022
									A070504	Premature Discharge of Battery	28/06/2022	28/06/2022
							F1905	Device Revision or Replacement	A0705	Battery Problem	28/06/2022	28/06/2022
									A070504	Premature Discharge of Battery	28/06/2022	28/06/2022
					CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem	28/06/2022	28/06/2022
									A070504	Premature Discharge of Battery	28/06/2022	28/06/2022
							F08	Hospitalization or Prolonged Hospitalization	A0705	Battery Problem	28/06/2022	28/06/2022
									A070504	Premature Discharge of Battery	28/06/2022	28/06/2022
							F1905	Device Revision or Replacement	A0705	Battery Problem	28/06/2022	28/06/2022
									A070504	Premature Discharge of Battery	28/06/2022	28/06/2022
					CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem	28/06/2022	28/06/2022
									A070504	Premature Discharge of Battery	28/06/2022	28/06/2022
							F08	Hospitalization or Prolonged Hospitalization	A0705	Battery Problem	28/06/2022	28/06/2022
									A070504	Premature Discharge of Battery	28/06/2022	28/06/2022
							F1905	Device Revision or Replacement	A0705	Battery Problem	28/06/2022	28/06/2022
									A070504	Premature Discharge of Battery	28/06/2022	28/06/2022
					CT233	Surgical	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem	28/06/2022	28/06/2022
									A070504	Premature Discharge of Battery	28/06/2022	28/06/2022
							F08	Hospitalization or Prolonged Hospitalization	A0705	Battery Problem	28/06/2022	28/06/2022
									A070504	Premature Discharge of Battery	28/06/2022	28/06/2022
							F1905	Device Revision or Replacement	A0705	Battery Problem	28/06/2022	28/06/2022
									A070504	Premature Discharge of Battery	28/06/2022	28/06/2022
					CT2406	Implantable	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem	28/06/2022	28/06/2022
									A070504	Premature Discharge of Battery	28/06/2022	28/06/2022
							F08	Hospitalization or Prolonged Hospitalization	A0705	Battery Problem	28/06/2022	28/06/2022
									A070504	Premature Discharge of Battery	28/06/2022	28/06/2022
							F1905	Device Revision or Replacement	A0705	Battery Problem	28/06/2022	28/06/2022
									A070504	Premature Discharge of Battery	28/06/2022	28/06/2022
					CT2407	Active implantable	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem	28/06/2022	28/06/2022
									A070504	Premature Discharge of Battery	28/06/2022	28/06/2022
							F08	Hospitalization or Prolonged Hospitalization	A0705	Battery Problem	28/06/2022	28/06/2022
									A070504	Premature Discharge of Battery	28/06/2022	28/06/2022
							F1905	Device Revision or Replacement	A0705	Battery Problem	28/06/2022	28/06/2022
									A070504	Premature Discharge of Battery	28/06/2022	28/06/2022
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem	28/06/2022	28/06/2022
									A070504	Premature Discharge of Battery	28/06/2022	28/06/2022

							F08	Hospitalization or Prolonged Hospitalization	A0705	Battery Problem	28/06/2022	28/06/2022
									A070504	Premature Discharge of Battery	28/06/2022	28/06/2022
							F1905	Device Revision or Replacement	A0705	Battery Problem	28/06/2022	28/06/2022
									A070504	Premature Discharge of Battery	28/06/2022	28/06/2022
					CT2727	Active implant devices/ancillaries	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem	28/06/2022	28/06/2022
									A070504	Premature Discharge of Battery	28/06/2022	28/06/2022
							F08	Hospitalization or Prolonged Hospitalization	A0705	Battery Problem	28/06/2022	28/06/2022
									A070504	Premature Discharge of Battery	28/06/2022	28/06/2022
							F1905	Device Revision or Replacement	A0705	Battery Problem	28/06/2022	28/06/2022
									A070504	Premature Discharge of Battery	28/06/2022	28/06/2022
					CT2750	Implantable cardiac electrical stimulators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem	28/06/2022	28/06/2022
									A070504	Premature Discharge of Battery	28/06/2022	28/06/2022
							F08	Hospitalization or Prolonged Hospitalization	A0705	Battery Problem	28/06/2022	28/06/2022
									A070504	Premature Discharge of Battery	28/06/2022	28/06/2022
							F1905	Device Revision or Replacement	A0705	Battery Problem	28/06/2022	28/06/2022
									A070504	Premature Discharge of Battery	28/06/2022	28/06/2022
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem	28/06/2022	28/06/2022
									A070504	Premature Discharge of Battery	28/06/2022	28/06/2022
							F08	Hospitalization or Prolonged Hospitalization	A0705	Battery Problem	28/06/2022	28/06/2022
									A070504	Premature Discharge of Battery	28/06/2022	28/06/2022
							F1905	Device Revision or Replacement	A0705	Battery Problem	28/06/2022	28/06/2022
									A070504	Premature Discharge of Battery	28/06/2022	28/06/2022
					CT2985	Active	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem	28/06/2022	28/06/2022
									A070504	Premature Discharge of Battery	28/06/2022	28/06/2022
							F08	Hospitalization or Prolonged Hospitalization	A0705	Battery Problem	28/06/2022	28/06/2022
									A070504	Premature Discharge of Battery	28/06/2022	28/06/2022
							F1905	Device Revision or Replacement	A0705	Battery Problem	28/06/2022	28/06/2022
									A070504	Premature Discharge of Battery	28/06/2022	28/06/2022
					CT3001	Implantable metal	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem	28/06/2022	28/06/2022
									A070504	Premature Discharge of Battery	28/06/2022	28/06/2022
							F08	Hospitalization or Prolonged Hospitalization	A0705	Battery Problem	28/06/2022	28/06/2022
									A070504	Premature Discharge of Battery	28/06/2022	28/06/2022
							F1905	Device Revision or Replacement	A0705	Battery Problem	28/06/2022	28/06/2022
									A070504	Premature Discharge of Battery	28/06/2022	28/06/2022
					CT3002	Implantable synthetic polymer	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem	28/06/2022	28/06/2022
									A070504	Premature Discharge of Battery	28/06/2022	28/06/2022
							F08	Hospitalization or Prolonged Hospitalization	A0705	Battery Problem	28/06/2022	28/06/2022
									A070504	Premature Discharge of Battery	28/06/2022	28/06/2022
							F1905	Device Revision or Replacement	A0705	Battery Problem	28/06/2022	28/06/2022
									A070504	Premature Discharge of Battery	28/06/2022	28/06/2022
					CT3006	Implantable inorganic material	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem	28/06/2022	28/06/2022
									A070504	Premature Discharge of Battery	28/06/2022	28/06/2022
							F08	Hospitalization or Prolonged Hospitalization	A0705	Battery Problem	28/06/2022	28/06/2022
									A070504	Premature Discharge of Battery	28/06/2022	28/06/2022
							F1905	Device Revision or Replacement	A0705	Battery Problem	28/06/2022	28/06/2022
									A070504	Premature Discharge of Battery	28/06/2022	28/06/2022
					CT314	Electrically powered	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem	28/06/2022	28/06/2022
									A070504	Premature Discharge of Battery	28/06/2022	28/06/2022
							F08	Hospitalization or Prolonged Hospitalization	A0705	Battery Problem	28/06/2022	28/06/2022
									A070504	Premature Discharge of Battery	28/06/2022	28/06/2022
							F1905	Device Revision or Replacement	A0705	Battery Problem	28/06/2022	28/06/2022
									A070504	Premature Discharge of Battery	28/06/2022	28/06/2022
					CT321	Long-term surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem	28/06/2022	28/06/2022
									A070504	Premature Discharge of Battery	28/06/2022	28/06/2022
							F08	Hospitalization or Prolonged Hospitalization	A0705	Battery Problem	28/06/2022	28/06/2022
									A070504	Premature Discharge of Battery	28/06/2022	28/06/2022
							F1905	Device Revision or Replacement	A0705	Battery Problem	28/06/2022	28/06/2022
									A070504	Premature Discharge of Battery	28/06/2022	28/06/2022
					CT449	Generators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem	28/06/2022	28/06/2022
									A070504	Premature Discharge of Battery	28/06/2022	28/06/2022
							F08	Hospitalization or Prolonged Hospitalization	A0705	Battery Problem	28/06/2022	28/06/2022
									A070504	Premature Discharge of Battery	28/06/2022	28/06/2022
							F1905	Device Revision or Replacement	A0705	Battery Problem	28/06/2022	28/06/2022
									A070504	Premature Discharge of Battery	28/06/2022	28/06/2022
					CT981	Single use	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem	28/06/2022	28/06/2022
									A070504	Premature Discharge of Battery	28/06/2022	28/06/2022
							F08	Hospitalization or Prolonged Hospitalization	A0705	Battery Problem	28/06/2022	28/06/2022
									A070504	Premature Discharge of Battery	28/06/2022	28/06/2022
							F1905	Device Revision or Replacement	A0705	Battery Problem	28/06/2022	28/06/2022
									A070504	Premature Discharge of Battery	28/06/2022	28/06/2022
					CT983	Surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem	28/06/2022	28/06/2022
									A070504	Premature Discharge of Battery	28/06/2022	28/06/2022
							F08	Hospitalization or Prolonged Hospitalization	A0705	Battery Problem	28/06/2022	28/06/2022
									A070504	Premature Discharge of Battery	28/06/2022	28/06/2022
							F1905	Device Revision or Replacement	A0705	Battery Problem	28/06/2022	28/06/2022
									A070504	Premature Discharge of Battery	28/06/2022	28/06/2022
6545	2021	All other reportable incidents	35223	Endocardial/interventricular septal pacing lead	CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A0401	Break		
									A070909	Oversensing		
							F26	No Health Consequences or Impact	A0401	Break		
									A070909	Oversensing		
6546	2022	All other reportable incidents	35223	Endocardial/interventricular septal pacing lead	CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070102	High Capture Threshold	18/03/2022	18/03/2022
									A070909	Oversensing	18/03/2022	18/03/2022
									A072201	High Impedance	18/03/2022	18/03/2022
									A090801	Signal Artifact/Noise	18/03/2022	18/03/2022
							F08	Hospitalization or Prolonged Hospitalization	A070102	High Capture Threshold	18/03/2022	18/03/2022
									A070909	Oversensing	18/03/2022	18/03/2022
									A072201	High Impedance	18/03/2022	18/03/2022
									A090801	Signal Artifact/Noise	18/03/2022	18/03/2022

							F12	Serious Injury/ Illness/ Impairment	A070102	High Capture Threshold	18/03/2022	18/03/2022
									A070909	Oversensing	18/03/2022	18/03/2022
									A072201	High Impedance	18/03/2022	18/03/2022
									A090801	Signal Artifact/Noise	18/03/2022	18/03/2022
							F1905	Device Revision or Replacement	A070102	High Capture Threshold	18/03/2022	18/03/2022
									A070909	Oversensing	18/03/2022	18/03/2022
									A072201	High Impedance	18/03/2022	18/03/2022
									A090801	Signal Artifact/Noise	18/03/2022	18/03/2022
6547	2022	All other reportable incidents	36241	Pacemaker/defibrillator lead	CT1684	Cardiac pacemakers and associated devices	E2103	Device Overstimulation of Tissue	A0203	Defective Device	12/05/2022	12/05/2022
									A071209	Extracardiac Stimulation	12/05/2022	12/05/2022
									A071302	Inappropriate/Inadequate Shock/Stimulation	12/05/2022	12/05/2022
									A072202	Low Impedance	12/05/2022	12/05/2022
									A090801	Signal Artifact/Noise	12/05/2022	12/05/2022
							E2104	Electric Shock	A0203	Defective Device	12/05/2022	12/05/2022
									A071209	Extracardiac Stimulation	12/05/2022	12/05/2022
									A071302	Inappropriate/Inadequate Shock/Stimulation	12/05/2022	12/05/2022
									A072202	Low Impedance	12/05/2022	12/05/2022
									A090801	Signal Artifact/Noise	12/05/2022	12/05/2022
							E2403	No Clinical Signs, Symptoms or Conditions	A0203	Defective Device	12/05/2022	12/05/2022
									A071209	Extracardiac Stimulation	12/05/2022	12/05/2022
									A071302	Inappropriate/Inadequate Shock/Stimulation	12/05/2022	12/05/2022
									A072202	Low Impedance	12/05/2022	12/05/2022
									A090801	Signal Artifact/Noise	12/05/2022	12/05/2022
							F08	Hospitalization or Prolonged Hospitalization	A0203	Defective Device	12/05/2022	12/05/2022
									A071209	Extracardiac Stimulation	12/05/2022	12/05/2022
									A071302	Inappropriate/Inadequate Shock/Stimulation	12/05/2022	12/05/2022
									A072202	Low Impedance	12/05/2022	12/05/2022
									A090801	Signal Artifact/Noise	12/05/2022	12/05/2022
							F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A0203	Defective Device	12/05/2022	12/05/2022
									A071209	Extracardiac Stimulation	12/05/2022	12/05/2022
									A071302	Inappropriate/Inadequate Shock/Stimulation	12/05/2022	12/05/2022
									A072202	Low Impedance	12/05/2022	12/05/2022
									A090801	Signal Artifact/Noise	12/05/2022	12/05/2022
							F1905	Device Revision or Replacement	A0203	Defective Device	12/05/2022	12/05/2022
									A071209	Extracardiac Stimulation	12/05/2022	12/05/2022
									A071302	Inappropriate/Inadequate Shock/Stimulation	12/05/2022	12/05/2022
									A072202	Low Impedance	12/05/2022	12/05/2022
									A090801	Signal Artifact/Noise	12/05/2022	12/05/2022
6548	2021	All other reportable incidents	35852	Single-chamber implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A0908	Incorrect, Inadequate or Imprecise Result or Readings	15/09/2021	15/09/2021
							F08	Hospitalization or Prolonged Hospitalization	A0908	Incorrect, Inadequate or Imprecise Result or Readings	15/09/2021	15/09/2021
							F1905	Device Revision or Replacement	A0908	Incorrect, Inadequate or Imprecise Result or Readings	15/09/2021	15/09/2021
6549	2021	All other reportable incidents	36241	Pacemaker/defibrillator lead	CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070909	Oversensing		
							F26	No Health Consequences or Impact	A070909	Oversensing		
6550	2021	All other reportable incidents	17882	Physiologic-monitoring defibrillation system	CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070803	Failure to Power Up		
							F27	Problem Identified During Non-Clinical Procedure	A070803	Failure to Power Up		
6551	2021	All other reportable incidents	35852	Single-chamber implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E2104	Electric Shock	A070909	Oversensing	18/02/2021	18/02/2021
									A071302	Inappropriate/Inadequate Shock/Stimulation	18/02/2021	18/02/2021
							E2403	No Clinical Signs, Symptoms or Conditions	A070909	Oversensing	18/02/2021	18/02/2021
									A071302	Inappropriate/Inadequate Shock/Stimulation	18/02/2021	18/02/2021
							F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A070909	Oversensing	18/02/2021	18/02/2021
									A071302	Inappropriate/Inadequate Shock/Stimulation	18/02/2021	18/02/2021
							F19	Surgical Intervention	A070909	Oversensing	18/02/2021	18/02/2021
									A071302	Inappropriate/Inadequate Shock/Stimulation	18/02/2021	18/02/2021
							F1903	Device Explantation	A070909	Oversensing	18/02/2021	18/02/2021
									A071302	Inappropriate/Inadequate Shock/Stimulation	18/02/2021	18/02/2021
6552	2021	All other reportable incidents	47265	Dual-chamber implantable pacemaker, rate-responsive	CT141	Electrodes/Leads and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	24/09/2021	24/09/2021
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	24/09/2021	24/09/2021
							F1903	Device Explantation	A070504	Premature Discharge of Battery	24/09/2021	24/09/2021
					CT145	Cardiovascular devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	24/09/2021	24/09/2021
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	24/09/2021	24/09/2021
							F1903	Device Explantation	A070504	Premature Discharge of Battery	24/09/2021	24/09/2021
					CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	24/09/2021	24/09/2021
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	24/09/2021	24/09/2021
							F1903	Device Explantation	A070504	Premature Discharge of Battery	24/09/2021	24/09/2021
					CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	24/09/2021	24/09/2021
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	24/09/2021	24/09/2021
							F1903	Device Explantation	A070504	Premature Discharge of Battery	24/09/2021	24/09/2021
					CT233	Surgical	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	24/09/2021	24/09/2021
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	24/09/2021	24/09/2021
							F1903	Device Explantation	A070504	Premature Discharge of Battery	24/09/2021	24/09/2021
					CT2406	Implantable	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	24/09/2021	24/09/2021
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	24/09/2021	24/09/2021
							F1903	Device Explantation	A070504	Premature Discharge of Battery	24/09/2021	24/09/2021
					CT2407	Active implantable	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	24/09/2021	24/09/2021
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	24/09/2021	24/09/2021
							F1903	Device Explantation	A070504	Premature Discharge of Battery	24/09/2021	24/09/2021
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	24/09/2021	24/09/2021
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	24/09/2021	24/09/2021
							F1903	Device Explantation	A070504	Premature Discharge of Battery	24/09/2021	24/09/2021
					CT2727	Active implant devices/ancillaries	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	24/09/2021	24/09/2021
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	24/09/2021	24/09/2021
							F1903	Device Explantation	A070504	Premature Discharge of Battery	24/09/2021	24/09/2021
					CT2750	Implantable cardiac electrical stimulators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	24/09/2021	24/09/2021
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	24/09/2021	24/09/2021
							F1903	Device Explantation	A070504	Premature Discharge of Battery	24/09/2021	24/09/2021

					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	24/09/2021	24/09/2021
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	24/09/2021	24/09/2021
							F1903	Device Explantation	A070504	Premature Discharge of Battery	24/09/2021	24/09/2021
					CT2985	Active	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	24/09/2021	24/09/2021
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	24/09/2021	24/09/2021
							F1903	Device Explantation	A070504	Premature Discharge of Battery	24/09/2021	24/09/2021
					CT3001	Implantable metal	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	24/09/2021	24/09/2021
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	24/09/2021	24/09/2021
							F1903	Device Explantation	A070504	Premature Discharge of Battery	24/09/2021	24/09/2021
					CT3002	Implantable synthetic polymer	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	24/09/2021	24/09/2021
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	24/09/2021	24/09/2021
							F1903	Device Explantation	A070504	Premature Discharge of Battery	24/09/2021	24/09/2021
					CT3006	Implantable inorganic material	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	24/09/2021	24/09/2021
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	24/09/2021	24/09/2021
							F1903	Device Explantation	A070504	Premature Discharge of Battery	24/09/2021	24/09/2021
					CT314	Electrically powered	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	24/09/2021	24/09/2021
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	24/09/2021	24/09/2021
							F1903	Device Explantation	A070504	Premature Discharge of Battery	24/09/2021	24/09/2021
					CT321	Long-term surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	24/09/2021	24/09/2021
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	24/09/2021	24/09/2021
							F1903	Device Explantation	A070504	Premature Discharge of Battery	24/09/2021	24/09/2021
					CT449	Generators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	24/09/2021	24/09/2021
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	24/09/2021	24/09/2021
							F1903	Device Explantation	A070504	Premature Discharge of Battery	24/09/2021	24/09/2021
					CT981	Single use	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	24/09/2021	24/09/2021
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	24/09/2021	24/09/2021
							F1903	Device Explantation	A070504	Premature Discharge of Battery	24/09/2021	24/09/2021
					CT983	Surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	24/09/2021	24/09/2021
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	24/09/2021	24/09/2021
							F1903	Device Explantation	A070504	Premature Discharge of Battery	24/09/2021	24/09/2021
6553	2022	All other reportable incidents	35223	Endocardial/interventricular septal pacing lead	CT1684	Cardiac pacemakers and associated devices	E060104	Bradycardia	A070102	High Capture Threshold		
									A0722	Impedance Problem		
							E2007	Fall	A070102	High Capture Threshold		
									A0722	Impedance Problem		
							E2403	No Clinical Signs, Symptoms or Conditions	A070102	High Capture Threshold		
									A0722	Impedance Problem		
							F12	Serious Injury/ Illness/ Impairment	A070102	High Capture Threshold		
									A0722	Impedance Problem		
							F23	Unexpected Medical Intervention	A070102	High Capture Threshold		
									A0722	Impedance Problem		
6554	2021	All other reportable incidents	47267	Single-chamber implantable pacemaker, rate-responsive	CT141	Electrodes/Leads and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	27/10/2021	27/10/2021
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	27/10/2021	27/10/2021
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	27/10/2021	27/10/2021
					CT145	Cardiovascular devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	27/10/2021	27/10/2021
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	27/10/2021	27/10/2021
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	27/10/2021	27/10/2021
					CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	27/10/2021	27/10/2021
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	27/10/2021	27/10/2021
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	27/10/2021	27/10/2021
					CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	27/10/2021	27/10/2021
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	27/10/2021	27/10/2021
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	27/10/2021	27/10/2021
					CT233	Surgical	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	27/10/2021	27/10/2021
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	27/10/2021	27/10/2021
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	27/10/2021	27/10/2021
					CT2406	Implantable	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	27/10/2021	27/10/2021
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	27/10/2021	27/10/2021
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	27/10/2021	27/10/2021
					CT2407	Active implantable	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	27/10/2021	27/10/2021
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	27/10/2021	27/10/2021
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	27/10/2021	27/10/2021
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	27/10/2021	27/10/2021
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	27/10/2021	27/10/2021
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	27/10/2021	27/10/2021
					CT2727	Active implant devices/ancillaries	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	27/10/2021	27/10/2021
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	27/10/2021	27/10/2021
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	27/10/2021	27/10/2021
					CT2750	Implantable cardiac electrical stimulators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	27/10/2021	27/10/2021
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	27/10/2021	27/10/2021
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	27/10/2021	27/10/2021
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	27/10/2021	27/10/2021
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	27/10/2021	27/10/2021
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	27/10/2021	27/10/2021
					CT2985	Active	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	27/10/2021	27/10/2021
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	27/10/2021	27/10/2021
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	27/10/2021	27/10/2021
					CT3001	Implantable metal	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	27/10/2021	27/10/2021
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	27/10/2021	27/10/2021
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	27/10/2021	27/10/2021
					CT3002	Implantable synthetic polymer	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	27/10/2021	27/10/2021
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	27/10/2021	27/10/2021
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	27/10/2021	27/10/2021
					CT3006	Implantable inorganic material	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	27/10/2021	27/10/2021
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	27/10/2021	27/10/2021
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	27/10/2021	27/10/2021
					CT314	Electrically powered	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	27/10/2021	27/10/2021
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	27/10/2021	27/10/2021
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	27/10/2021	27/10/2021

					CT321	Long-term surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	27/10/2021	27/10/2021
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	27/10/2021	27/10/2021
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	27/10/2021	27/10/2021
					CT449	Generators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	27/10/2021	27/10/2021
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	27/10/2021	27/10/2021
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	27/10/2021	27/10/2021
					CT981	Single use	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	27/10/2021	27/10/2021
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	27/10/2021	27/10/2021
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	27/10/2021	27/10/2021
					CT983	Surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	27/10/2021	27/10/2021
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	27/10/2021	27/10/2021
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	27/10/2021	27/10/2021
6555	2021	All other reportable incidents	35852	Single-chamber implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem	20/01/2022	20/01/2022
									A070504	Premature Discharge of Battery	20/01/2022	20/01/2022
							F08	Hospitalization or Prolonged Hospitalization	A0705	Battery Problem	20/01/2022	20/01/2022
									A070504	Premature Discharge of Battery	20/01/2022	20/01/2022
							F1905	Device Revision or Replacement	A0705	Battery Problem	20/01/2022	20/01/2022
									A070504	Premature Discharge of Battery	20/01/2022	20/01/2022
							F23	Unexpected Medical Intervention	A0705	Battery Problem	20/01/2022	20/01/2022
									A070504	Premature Discharge of Battery	20/01/2022	20/01/2022
6556	2021	All other reportable incidents	37265	Dual-chamber implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	22/09/2021	22/09/2021
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	22/09/2021	22/09/2021
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	22/09/2021	22/09/2021
6557	2021	All other reportable incidents	35852	Single-chamber implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem	04/10/2021	04/10/2021
									A070504	Premature Discharge of Battery	04/10/2021	04/10/2021
							F08	Hospitalization or Prolonged Hospitalization	A0705	Battery Problem	04/10/2021	04/10/2021
									A070504	Premature Discharge of Battery	04/10/2021	04/10/2021
							F1905	Device Revision or Replacement	A0705	Battery Problem	04/10/2021	04/10/2021
									A070504	Premature Discharge of Battery	04/10/2021	04/10/2021
6558	2021	All other reportable incidents	35852	Single-chamber implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem	03/09/2021	03/09/2021
									A070504	Premature Discharge of Battery	03/09/2021	03/09/2021
							F08	Hospitalization or Prolonged Hospitalization	A0705	Battery Problem	03/09/2021	03/09/2021
									A070504	Premature Discharge of Battery	03/09/2021	03/09/2021
							F1905	Device Revision or Replacement	A0705	Battery Problem	03/09/2021	03/09/2021
									A070504	Premature Discharge of Battery	03/09/2021	03/09/2021
6559	2021	All other reportable incidents	35852	Single-chamber implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A0711	Interrogation Problem	10/09/2021	10/09/2021
							F26	No Health Consequences or Impact	A0711	Interrogation Problem	10/09/2021	10/09/2021
6560	2021	All other reportable incidents	35852	Single-chamber implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem	27/10/2021	27/10/2021
									A070504	Premature Discharge of Battery	27/10/2021	27/10/2021
							F08	Hospitalization or Prolonged Hospitalization	A0705	Battery Problem	27/10/2021	27/10/2021
									A070504	Premature Discharge of Battery	27/10/2021	27/10/2021
							F1905	Device Revision or Replacement	A0705	Battery Problem	27/10/2021	27/10/2021
									A070504	Premature Discharge of Battery	27/10/2021	27/10/2021
6561	2021	All other reportable incidents	47265	Dual-chamber implantable pacemaker, rate-responsive	CT141	Electrodes/Leads and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	01/09/2021	01/09/2021
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	01/09/2021	01/09/2021
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	01/09/2021	01/09/2021
					CT145	Cardiovascular devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	01/09/2021	01/09/2021
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	01/09/2021	01/09/2021
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	01/09/2021	01/09/2021
					CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	01/09/2021	01/09/2021
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	01/09/2021	01/09/2021
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	01/09/2021	01/09/2021
					CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	01/09/2021	01/09/2021
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	01/09/2021	01/09/2021
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	01/09/2021	01/09/2021
					CT233	Surgical	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	01/09/2021	01/09/2021
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	01/09/2021	01/09/2021
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	01/09/2021	01/09/2021
					CT2406	Implantable	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	01/09/2021	01/09/2021
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	01/09/2021	01/09/2021
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	01/09/2021	01/09/2021
					CT2407	Active implantable	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	01/09/2021	01/09/2021
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	01/09/2021	01/09/2021
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	01/09/2021	01/09/2021
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	01/09/2021	01/09/2021
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	01/09/2021	01/09/2021
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	01/09/2021	01/09/2021
					CT2727	Active implant devices/ancillaries	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	01/09/2021	01/09/2021
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	01/09/2021	01/09/2021
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	01/09/2021	01/09/2021
					CT2750	Implantable cardiac electrical stimulators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	01/09/2021	01/09/2021
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	01/09/2021	01/09/2021
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	01/09/2021	01/09/2021
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	01/09/2021	01/09/2021
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	01/09/2021	01/09/2021
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	01/09/2021	01/09/2021
					CT2985	Active	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	01/09/2021	01/09/2021
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	01/09/2021	01/09/2021
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	01/09/2021	01/09/2021
					CT3001	Implantable metal	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	01/09/2021	01/09/2021
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	01/09/2021	01/09/2021
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	01/09/2021	01/09/2021
					CT3002	Implantable synthetic polymer	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	01/09/2021	01/09/2021
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	01/09/2021	01/09/2021
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	01/09/2021	01/09/2021
					CT3006	Implantable inorganic material	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	01/09/2021	01/09/2021
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	01/09/2021	01/09/2021
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	01/09/2021	01/09/2021

					CT314	Electrically powered	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	01/09/2021	01/09/2021
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	01/09/2021	01/09/2021
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	01/09/2021	01/09/2021
					CT321	Long-term surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	01/09/2021	01/09/2021
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	01/09/2021	01/09/2021
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	01/09/2021	01/09/2021
					CT449	Generators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	01/09/2021	01/09/2021
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	01/09/2021	01/09/2021
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	01/09/2021	01/09/2021
					CT981	Single use	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	01/09/2021	01/09/2021
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	01/09/2021	01/09/2021
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	01/09/2021	01/09/2021
					CT983	Surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	01/09/2021	01/09/2021
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	01/09/2021	01/09/2021
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	01/09/2021	01/09/2021
6562	2022	All other reportable incidents	35223	Endocardial/interventricular septal pacing lead	CT1684	Cardiac pacemakers and associated devices	E0119	Loss of Consciousness	A07	Electrical /Electronic Property Problem		
									A070101	Failure to Capture		
									A072201	High Impedance		
							E2403	No Clinical Signs, Symptoms or Conditions	A07	Electrical /Electronic Property Problem		
									A070101	Failure to Capture		
									A072201	High Impedance		
							F12	Serious Injury/ Illness/ Impairment	A07	Electrical /Electronic Property Problem		
									A070101	Failure to Capture		
									A072201	High Impedance		
							F23	Unexpected Medical Intervention	A07	Electrical /Electronic Property Problem		
									A070101	Failure to Capture		
									A072201	High Impedance		
6563	2022	All other reportable incidents	60789	Intracardiac pacemaker	CT130	Catheters and associated devices	E0611	Heart Failure/Congestive Heart Failure	A27	Appropriate Device Problem Term/Code not Available		
							E0623	Unspecified Heart Problem	A27	Appropriate Device Problem Term/Code not Available		
							E2330	Pain	A27	Appropriate Device Problem Term/Code not Available		
							E2403	No Clinical Signs, Symptoms or Conditions	A27	Appropriate Device Problem Term/Code not Available		
							F12	Serious Injury/ Illness/ Impairment	A27	Appropriate Device Problem Term/Code not Available		
					CT145	Cardiovascular devices	E0611	Heart Failure/Congestive Heart Failure	A27	Appropriate Device Problem Term/Code not Available		
							E0623	Unspecified Heart Problem	A27	Appropriate Device Problem Term/Code not Available		
							E2330	Pain	A27	Appropriate Device Problem Term/Code not Available		
							E2403	No Clinical Signs, Symptoms or Conditions	A27	Appropriate Device Problem Term/Code not Available		
							F12	Serious Injury/ Illness/ Impairment	A27	Appropriate Device Problem Term/Code not Available		
					CT1684	Cardiac pacemakers and associated devices	E0611	Heart Failure/Congestive Heart Failure	A27	Appropriate Device Problem Term/Code not Available		
							E0623	Unspecified Heart Problem	A27	Appropriate Device Problem Term/Code not Available		
							E2330	Pain	A27	Appropriate Device Problem Term/Code not Available		
							E2403	No Clinical Signs, Symptoms or Conditions	A27	Appropriate Device Problem Term/Code not Available		
							F12	Serious Injury/ Illness/ Impairment	A27	Appropriate Device Problem Term/Code not Available		
					CT1685	Cardiac pacemakers	E0611	Heart Failure/Congestive Heart Failure	A27	Appropriate Device Problem Term/Code not Available		
							E0623	Unspecified Heart Problem	A27	Appropriate Device Problem Term/Code not Available		
							E2330	Pain	A27	Appropriate Device Problem Term/Code not Available		
							E2403	No Clinical Signs, Symptoms or Conditions	A27	Appropriate Device Problem Term/Code not Available		
							F12	Serious Injury/ Illness/ Impairment	A27	Appropriate Device Problem Term/Code not Available		
					CT233	Surgical	E0611	Heart Failure/Congestive Heart Failure	A27	Appropriate Device Problem Term/Code not Available		
							E0623	Unspecified Heart Problem	A27	Appropriate Device Problem Term/Code not Available		
							E2330	Pain	A27	Appropriate Device Problem Term/Code not Available		
							E2403	No Clinical Signs, Symptoms or Conditions	A27	Appropriate Device Problem Term/Code not Available		
							F12	Serious Injury/ Illness/ Impairment	A27	Appropriate Device Problem Term/Code not Available		
					CT2406	Implantable	E0611	Heart Failure/Congestive Heart Failure	A27	Appropriate Device Problem Term/Code not Available		
							E0623	Unspecified Heart Problem	A27	Appropriate Device Problem Term/Code not Available		
							E2330	Pain	A27	Appropriate Device Problem Term/Code not Available		
							E2403	No Clinical Signs, Symptoms or Conditions	A27	Appropriate Device Problem Term/Code not Available		
							F12	Serious Injury/ Illness/ Impairment	A27	Appropriate Device Problem Term/Code not Available		
					CT2407	Active implantable	E0611	Heart Failure/Congestive Heart Failure	A27	Appropriate Device Problem Term/Code not Available		
							E0623	Unspecified Heart Problem	A27	Appropriate Device Problem Term/Code not Available		
							E2330	Pain	A27	Appropriate Device Problem Term/Code not Available		
							E2403	No Clinical Signs, Symptoms or Conditions	A27	Appropriate Device Problem Term/Code not Available		

							F12	Serious Injury/ Illness/ Impairment	A27	Appropriate Device Problem Term/Code not Available		
					CT257	Implantable cardiac pacemakers	E0611	Heart Failure/Congestive Heart Failure	A27	Appropriate Device Problem Term/Code not Available		
							E0623	Unspecified Heart Problem	A27	Appropriate Device Problem Term/Code not Available		
							E2330	Pain	A27	Appropriate Device Problem Term/Code not Available		
							E2403	No Clinical Signs, Symptoms or Conditions	A27	Appropriate Device Problem Term/Code not Available		
							F12	Serious Injury/ Illness/ Impairment	A27	Appropriate Device Problem Term/Code not Available		
					CT2727	Active implant devices/ancillaries	E0611	Heart Failure/Congestive Heart Failure	A27	Appropriate Device Problem Term/Code not Available		
							E0623	Unspecified Heart Problem	A27	Appropriate Device Problem Term/Code not Available		
							E2330	Pain	A27	Appropriate Device Problem Term/Code not Available		
							E2403	No Clinical Signs, Symptoms or Conditions	A27	Appropriate Device Problem Term/Code not Available		
							F12	Serious Injury/ Illness/ Impairment	A27	Appropriate Device Problem Term/Code not Available		
					CT2750	Implantable cardiac electrical stimulators and associated devices	E0611	Heart Failure/Congestive Heart Failure	A27	Appropriate Device Problem Term/Code not Available		
							E0623	Unspecified Heart Problem	A27	Appropriate Device Problem Term/Code not Available		
							E2330	Pain	A27	Appropriate Device Problem Term/Code not Available		
							E2403	No Clinical Signs, Symptoms or Conditions	A27	Appropriate Device Problem Term/Code not Available		
							F12	Serious Injury/ Illness/ Impairment	A27	Appropriate Device Problem Term/Code not Available		
					CT2818	Catheterization/Cannulation/Fluid path devices	E0611	Heart Failure/Congestive Heart Failure	A27	Appropriate Device Problem Term/Code not Available		
							E0623	Unspecified Heart Problem	A27	Appropriate Device Problem Term/Code not Available		
							E2330	Pain	A27	Appropriate Device Problem Term/Code not Available		
							E2403	No Clinical Signs, Symptoms or Conditions	A27	Appropriate Device Problem Term/Code not Available		
							F12	Serious Injury/ Illness/ Impairment	A27	Appropriate Device Problem Term/Code not Available		
					CT2866	Implantable cardiac electrical stimulators	E0611	Heart Failure/Congestive Heart Failure	A27	Appropriate Device Problem Term/Code not Available		
							E0623	Unspecified Heart Problem	A27	Appropriate Device Problem Term/Code not Available		
							E2330	Pain	A27	Appropriate Device Problem Term/Code not Available		
							E2403	No Clinical Signs, Symptoms or Conditions	A27	Appropriate Device Problem Term/Code not Available		
							F12	Serious Injury/ Illness/ Impairment	A27	Appropriate Device Problem Term/Code not Available		
					CT2985	Active	E0611	Heart Failure/Congestive Heart Failure	A27	Appropriate Device Problem Term/Code not Available		
							E0623	Unspecified Heart Problem	A27	Appropriate Device Problem Term/Code not Available		
							E2330	Pain	A27	Appropriate Device Problem Term/Code not Available		
							E2403	No Clinical Signs, Symptoms or Conditions	A27	Appropriate Device Problem Term/Code not Available		
							F12	Serious Injury/ Illness/ Impairment	A27	Appropriate Device Problem Term/Code not Available		
					CT3001	Implantable metal	E0611	Heart Failure/Congestive Heart Failure	A27	Appropriate Device Problem Term/Code not Available		
							E0623	Unspecified Heart Problem	A27	Appropriate Device Problem Term/Code not Available		
							E2330	Pain	A27	Appropriate Device Problem Term/Code not Available		
							E2403	No Clinical Signs, Symptoms or Conditions	A27	Appropriate Device Problem Term/Code not Available		
							F12	Serious Injury/ Illness/ Impairment	A27	Appropriate Device Problem Term/Code not Available		
					CT3002	Implantable synthetic polymer	E0611	Heart Failure/Congestive Heart Failure	A27	Appropriate Device Problem Term/Code not Available		
							E0623	Unspecified Heart Problem	A27	Appropriate Device Problem Term/Code not Available		
							E2330	Pain	A27	Appropriate Device Problem Term/Code not Available		
							E2403	No Clinical Signs, Symptoms or Conditions	A27	Appropriate Device Problem Term/Code not Available		
							F12	Serious Injury/ Illness/ Impairment	A27	Appropriate Device Problem Term/Code not Available		
					CT3006	Implantable inorganic material	E0611	Heart Failure/Congestive Heart Failure	A27	Appropriate Device Problem Term/Code not Available		
							E0623	Unspecified Heart Problem	A27	Appropriate Device Problem Term/Code not Available		
							E2330	Pain	A27	Appropriate Device Problem Term/Code not Available		
							E2403	No Clinical Signs, Symptoms or Conditions	A27	Appropriate Device Problem Term/Code not Available		
							F12	Serious Injury/ Illness/ Impairment	A27	Appropriate Device Problem Term/Code not Available		
					CT314	Electrically powered	E0611	Heart Failure/Congestive Heart Failure	A27	Appropriate Device Problem Term/Code not Available		
							E0623	Unspecified Heart Problem	A27	Appropriate Device Problem Term/Code not Available		
							E2330	Pain	A27	Appropriate Device Problem Term/Code not Available		
							E2403	No Clinical Signs, Symptoms or Conditions	A27	Appropriate Device Problem Term/Code not Available		

						F12	Serious Injury/ Illness/ Impairment	A27	Appropriate Device Problem Term/Code not Available		
					CT321	Long-term surgical invasive	E0611	Heart Failure/Congestive Heart Failure	A27	Appropriate Device Problem Term/Code not Available		
							E0623	Unspecified Heart Problem	A27	Appropriate Device Problem Term/Code not Available		
							E2330	Pain	A27	Appropriate Device Problem Term/Code not Available		
							E2403	No Clinical Signs, Symptoms or Conditions	A27	Appropriate Device Problem Term/Code not Available		
							F12	Serious Injury/ Illness/ Impairment	A27	Appropriate Device Problem Term/Code not Available		
					CT981	Single use	E0611	Heart Failure/Congestive Heart Failure	A27	Appropriate Device Problem Term/Code not Available		
							E0623	Unspecified Heart Problem	A27	Appropriate Device Problem Term/Code not Available		
							E2330	Pain	A27	Appropriate Device Problem Term/Code not Available		
							E2403	No Clinical Signs, Symptoms or Conditions	A27	Appropriate Device Problem Term/Code not Available		
							F12	Serious Injury/ Illness/ Impairment	A27	Appropriate Device Problem Term/Code not Available		
					CT983	Surgical invasive	E0611	Heart Failure/Congestive Heart Failure	A27	Appropriate Device Problem Term/Code not Available		
							E0623	Unspecified Heart Problem	A27	Appropriate Device Problem Term/Code not Available		
							E2330	Pain	A27	Appropriate Device Problem Term/Code not Available		
							E2403	No Clinical Signs, Symptoms or Conditions	A27	Appropriate Device Problem Term/Code not Available		
							F12	Serious Injury/ Illness/ Impairment	A27	Appropriate Device Problem Term/Code not Available		
6564	2021	All other reportable incidents	17882	Physiologic-monitoring defibrillation system	CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A1603	Failure of Device to Self-Test		
							F27	Problem Identified During Non-Clinical Procedure	A1603	Failure of Device to Self-Test		
6565	2022	All other reportable incidents	47263	Cardiac resynchronization therapy implantable pacemaker	CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A0908	Incorrect, Inadequate or Imprecise Result or Readings	07/04/2022	07/04/2022
							F08	Hospitalization or Prolonged Hospitalization	A0908	Incorrect, Inadequate or Imprecise Result or Readings	07/04/2022	07/04/2022
							F1905	Device Revision or Replacement	A0908	Incorrect, Inadequate or Imprecise Result or Readings	07/04/2022	07/04/2022
					CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A0908	Incorrect, Inadequate or Imprecise Result or Readings	07/04/2022	07/04/2022
							F08	Hospitalization or Prolonged Hospitalization	A0908	Incorrect, Inadequate or Imprecise Result or Readings	07/04/2022	07/04/2022
							F1905	Device Revision or Replacement	A0908	Incorrect, Inadequate or Imprecise Result or Readings	07/04/2022	07/04/2022
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A0908	Incorrect, Inadequate or Imprecise Result or Readings	07/04/2022	07/04/2022
							F08	Hospitalization or Prolonged Hospitalization	A0908	Incorrect, Inadequate or Imprecise Result or Readings	07/04/2022	07/04/2022
							F1905	Device Revision or Replacement	A0908	Incorrect, Inadequate or Imprecise Result or Readings	07/04/2022	07/04/2022
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A0908	Incorrect, Inadequate or Imprecise Result or Readings	07/04/2022	07/04/2022
							F08	Hospitalization or Prolonged Hospitalization	A0908	Incorrect, Inadequate or Imprecise Result or Readings	07/04/2022	07/04/2022
							F1905	Device Revision or Replacement	A0908	Incorrect, Inadequate or Imprecise Result or Readings	07/04/2022	07/04/2022
6566	2021	All other reportable incidents	35852	Single-chamber implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem	13/10/2021	13/10/2021
									A070504	Premature Discharge of Battery	13/10/2021	13/10/2021
							F08	Hospitalization or Prolonged Hospitalization	A0705	Battery Problem	13/10/2021	13/10/2021
									A070504	Premature Discharge of Battery	13/10/2021	13/10/2021
							F1905	Device Revision or Replacement	A0705	Battery Problem	13/10/2021	13/10/2021
									A070504	Premature Discharge of Battery	13/10/2021	13/10/2021
6567	2022	All other reportable incidents	35852	Single-chamber implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A0203	Defective Device	13/04/2022	13/04/2022
									A12	Connection Problem	13/04/2022	13/04/2022
							F26	No Health Consequences or Impact	A0203	Defective Device	13/04/2022	13/04/2022
									A12	Connection Problem	13/04/2022	13/04/2022
6568	2021	All other reportable incidents	37265	Dual-chamber implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	26/08/2021	26/08/2021
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery	26/08/2021	26/08/2021
6569	2021	All other reportable incidents	36241	Pacemaker/defibrillator lead	CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070102	High Capture Threshold		
									A072201	High Impedance		
							F26	No Health Consequences or Impact	A070102	High Capture Threshold		
									A072201	High Impedance		
6570	2021	All other reportable incidents	47265	Dual-chamber implantable pacemaker, rate-responsive	CT141	Electrodes/Leads and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	27/09/2021	27/09/2021
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	27/09/2021	27/09/2021
							F1903	Device Explantation	A070504	Premature Discharge of Battery	27/09/2021	27/09/2021
					CT145	Cardiovascular devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	27/09/2021	27/09/2021
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	27/09/2021	27/09/2021
							F1903	Device Explantation	A070504	Premature Discharge of Battery	27/09/2021	27/09/2021
					CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	27/09/2021	27/09/2021
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	27/09/2021	27/09/2021
							F1903	Device Explantation	A070504	Premature Discharge of Battery	27/09/2021	27/09/2021
					CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	27/09/2021	27/09/2021
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	27/09/2021	27/09/2021
							F1903	Device Explantation	A070504	Premature Discharge of Battery	27/09/2021	27/09/2021
					CT233	Surgical	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	27/09/2021	27/09/2021
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	27/09/2021	27/09/2021
							F1903	Device Explantation	A070504	Premature Discharge of Battery	27/09/2021	27/09/2021
					CT2406	Implantable	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	27/09/2021	27/09/2021
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	27/09/2021	27/09/2021
							F1903	Device Explantation	A070504	Premature Discharge of Battery	27/09/2021	27/09/2021
					CT2407	Active implantable	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	27/09/2021	27/09/2021
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	27/09/2021	27/09/2021

							F1903	Device Explantation	A070504	Premature Discharge of Battery	27/09/2021	27/09/2021
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	27/09/2021	27/09/2021
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	27/09/2021	27/09/2021
							F1903	Device Explantation	A070504	Premature Discharge of Battery	27/09/2021	27/09/2021
					CT2727	Active implant devices/ancillaries	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	27/09/2021	27/09/2021
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	27/09/2021	27/09/2021
							F1903	Device Explantation	A070504	Premature Discharge of Battery	27/09/2021	27/09/2021
					CT2750	Implantable cardiac electrical stimulators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	27/09/2021	27/09/2021
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	27/09/2021	27/09/2021
							F1903	Device Explantation	A070504	Premature Discharge of Battery	27/09/2021	27/09/2021
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	27/09/2021	27/09/2021
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	27/09/2021	27/09/2021
							F1903	Device Explantation	A070504	Premature Discharge of Battery	27/09/2021	27/09/2021
					CT2985	Active	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	27/09/2021	27/09/2021
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	27/09/2021	27/09/2021
							F1903	Device Explantation	A070504	Premature Discharge of Battery	27/09/2021	27/09/2021
					CT3001	Implantable metal	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	27/09/2021	27/09/2021
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	27/09/2021	27/09/2021
							F1903	Device Explantation	A070504	Premature Discharge of Battery	27/09/2021	27/09/2021
					CT3002	Implantable synthetic polymer	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	27/09/2021	27/09/2021
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	27/09/2021	27/09/2021
							F1903	Device Explantation	A070504	Premature Discharge of Battery	27/09/2021	27/09/2021
					CT3006	Implantable inorganic material	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	27/09/2021	27/09/2021
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	27/09/2021	27/09/2021
							F1903	Device Explantation	A070504	Premature Discharge of Battery	27/09/2021	27/09/2021
					CT314	Electrically powered	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	27/09/2021	27/09/2021
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	27/09/2021	27/09/2021
							F1903	Device Explantation	A070504	Premature Discharge of Battery	27/09/2021	27/09/2021
					CT321	Long-term surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	27/09/2021	27/09/2021
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	27/09/2021	27/09/2021
							F1903	Device Explantation	A070504	Premature Discharge of Battery	27/09/2021	27/09/2021
					CT449	Generators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	27/09/2021	27/09/2021
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	27/09/2021	27/09/2021
							F1903	Device Explantation	A070504	Premature Discharge of Battery	27/09/2021	27/09/2021
					CT981	Single use	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	27/09/2021	27/09/2021
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	27/09/2021	27/09/2021
							F1903	Device Explantation	A070504	Premature Discharge of Battery	27/09/2021	27/09/2021
					CT983	Surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	27/09/2021	27/09/2021
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	27/09/2021	27/09/2021
							F1903	Device Explantation	A070504	Premature Discharge of Battery	27/09/2021	27/09/2021
6571	2021	All other reportable incidents	47265	Dual-chamber implantable pacemaker, rate-responsive	CT141	Electrodes/Leads and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	02/09/2021	02/09/2021
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	02/09/2021	02/09/2021
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	02/09/2021	02/09/2021
					CT145	Cardiovascular devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	02/09/2021	02/09/2021
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	02/09/2021	02/09/2021
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	02/09/2021	02/09/2021
					CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	02/09/2021	02/09/2021
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	02/09/2021	02/09/2021
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	02/09/2021	02/09/2021
					CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	02/09/2021	02/09/2021
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	02/09/2021	02/09/2021
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	02/09/2021	02/09/2021
					CT233	Surgical	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	02/09/2021	02/09/2021
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	02/09/2021	02/09/2021
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	02/09/2021	02/09/2021
					CT2406	Implantable	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	02/09/2021	02/09/2021
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	02/09/2021	02/09/2021
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	02/09/2021	02/09/2021
					CT2407	Active implantable	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	02/09/2021	02/09/2021
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	02/09/2021	02/09/2021
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	02/09/2021	02/09/2021
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	02/09/2021	02/09/2021
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	02/09/2021	02/09/2021
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	02/09/2021	02/09/2021
					CT2727	Active implant devices/ancillaries	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	02/09/2021	02/09/2021
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	02/09/2021	02/09/2021
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	02/09/2021	02/09/2021
					CT2750	Implantable cardiac electrical stimulators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	02/09/2021	02/09/2021
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	02/09/2021	02/09/2021
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	02/09/2021	02/09/2021
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	02/09/2021	02/09/2021
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	02/09/2021	02/09/2021
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	02/09/2021	02/09/2021
					CT2985	Active	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	02/09/2021	02/09/2021
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	02/09/2021	02/09/2021
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	02/09/2021	02/09/2021
					CT3001	Implantable metal	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	02/09/2021	02/09/2021
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	02/09/2021	02/09/2021
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	02/09/2021	02/09/2021
					CT3002	Implantable synthetic polymer	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	02/09/2021	02/09/2021
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	02/09/2021	02/09/2021
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	02/09/2021	02/09/2021
					CT3006	Implantable inorganic material	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	02/09/2021	02/09/2021
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	02/09/2021	02/09/2021
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	02/09/2021	02/09/2021
					CT314	Electrically powered	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	02/09/2021	02/09/2021
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	02/09/2021	02/09/2021
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	02/09/2021	02/09/2021

					CT321	Long-term surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	02/09/2021	02/09/2021
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	02/09/2021	02/09/2021
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	02/09/2021	02/09/2021
					CT449	Generators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	02/09/2021	02/09/2021
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	02/09/2021	02/09/2021
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	02/09/2021	02/09/2021
					CT981	Single use	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	02/09/2021	02/09/2021
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	02/09/2021	02/09/2021
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	02/09/2021	02/09/2021
					CT983	Surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	02/09/2021	02/09/2021
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	02/09/2021	02/09/2021
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	02/09/2021	02/09/2021
6572	2022	All other reportable incidents	35223	Endocardial/interventricular septal pacing lead	CT1684	Cardiac pacemakers and associated devices	E060109	Tachycardia	A070910	Undersensing		
							E2403	No Clinical Signs, Symptoms or Conditions	A070910	Undersensing		
							F12	Serious Injury/ Illness/ Impairment	A070910	Undersensing		
6573	2022	All other reportable incidents	47265	Dual-chamber implantable pacemaker, rate-responsive	CT141	Electrodes/Leads and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A05	Mechanical Problem	22/04/2022	22/04/2022
									A070101	Failure to Capture	22/04/2022	22/04/2022
									A0709	Device Sensing Problem	22/04/2022	22/04/2022
									A0712	Pacing Problem	22/04/2022	22/04/2022
									A072201	High Impedance	22/04/2022	22/04/2022
							F26	No Health Consequences or Impact	A05	Mechanical Problem	22/04/2022	22/04/2022
									A070101	Failure to Capture	22/04/2022	22/04/2022
									A0709	Device Sensing Problem	22/04/2022	22/04/2022
									A0712	Pacing Problem	22/04/2022	22/04/2022
									A072201	High Impedance	22/04/2022	22/04/2022
					CT145	Cardiovascular devices	E2403	No Clinical Signs, Symptoms or Conditions	A05	Mechanical Problem	22/04/2022	22/04/2022
									A070101	Failure to Capture	22/04/2022	22/04/2022
									A0709	Device Sensing Problem	22/04/2022	22/04/2022
									A0712	Pacing Problem	22/04/2022	22/04/2022
									A072201	High Impedance	22/04/2022	22/04/2022
							F26	No Health Consequences or Impact	A05	Mechanical Problem	22/04/2022	22/04/2022
									A070101	Failure to Capture	22/04/2022	22/04/2022
									A0709	Device Sensing Problem	22/04/2022	22/04/2022
									A0712	Pacing Problem	22/04/2022	22/04/2022
									A072201	High Impedance	22/04/2022	22/04/2022
					CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A05	Mechanical Problem	22/04/2022	22/04/2022
									A070101	Failure to Capture	22/04/2022	22/04/2022
									A0709	Device Sensing Problem	22/04/2022	22/04/2022
									A0712	Pacing Problem	22/04/2022	22/04/2022
									A072201	High Impedance	22/04/2022	22/04/2022
							F26	No Health Consequences or Impact	A05	Mechanical Problem	22/04/2022	22/04/2022
									A070101	Failure to Capture	22/04/2022	22/04/2022
									A0709	Device Sensing Problem	22/04/2022	22/04/2022
									A0712	Pacing Problem	22/04/2022	22/04/2022
									A072201	High Impedance	22/04/2022	22/04/2022
					CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A05	Mechanical Problem	22/04/2022	22/04/2022
									A070101	Failure to Capture	22/04/2022	22/04/2022
									A0709	Device Sensing Problem	22/04/2022	22/04/2022
									A0712	Pacing Problem	22/04/2022	22/04/2022
									A072201	High Impedance	22/04/2022	22/04/2022
							F26	No Health Consequences or Impact	A05	Mechanical Problem	22/04/2022	22/04/2022
									A070101	Failure to Capture	22/04/2022	22/04/2022
									A0709	Device Sensing Problem	22/04/2022	22/04/2022
									A0712	Pacing Problem	22/04/2022	22/04/2022
									A072201	High Impedance	22/04/2022	22/04/2022
					CT233	Surgical	E2403	No Clinical Signs, Symptoms or Conditions	A05	Mechanical Problem	22/04/2022	22/04/2022
									A070101	Failure to Capture	22/04/2022	22/04/2022
									A0709	Device Sensing Problem	22/04/2022	22/04/2022
									A0712	Pacing Problem	22/04/2022	22/04/2022
									A072201	High Impedance	22/04/2022	22/04/2022
							F26	No Health Consequences or Impact	A05	Mechanical Problem	22/04/2022	22/04/2022
									A070101	Failure to Capture	22/04/2022	22/04/2022
									A0709	Device Sensing Problem	22/04/2022	22/04/2022
									A0712	Pacing Problem	22/04/2022	22/04/2022
									A072201	High Impedance	22/04/2022	22/04/2022
					CT2406	Implantable	E2403	No Clinical Signs, Symptoms or Conditions	A05	Mechanical Problem	22/04/2022	22/04/2022
									A070101	Failure to Capture	22/04/2022	22/04/2022
									A0709	Device Sensing Problem	22/04/2022	22/04/2022
									A0712	Pacing Problem	22/04/2022	22/04/2022
									A072201	High Impedance	22/04/2022	22/04/2022
							F26	No Health Consequences or Impact	A05	Mechanical Problem	22/04/2022	22/04/2022
									A070101	Failure to Capture	22/04/2022	22/04/2022
									A0709	Device Sensing Problem	22/04/2022	22/04/2022
									A0712	Pacing Problem	22/04/2022	22/04/2022
									A072201	High Impedance	22/04/2022	22/04/2022
					CT2407	Active implantable	E2403	No Clinical Signs, Symptoms or Conditions	A05	Mechanical Problem	22/04/2022	22/04/2022
									A070101	Failure to Capture	22/04/2022	22/04/2022
									A0709	Device Sensing Problem	22/04/2022	22/04/2022
									A0712	Pacing Problem	22/04/2022	22/04/2022
									A072201	High Impedance	22/04/2022	22/04/2022
							F26	No Health Consequences or Impact	A05	Mechanical Problem	22/04/2022	22/04/2022
									A070101	Failure to Capture	22/04/2022	22/04/2022
									A0709	Device Sensing Problem	22/04/2022	22/04/2022
									A0712	Pacing Problem	22/04/2022	22/04/2022
									A072201	High Impedance	22/04/2022	22/04/2022
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A05	Mechanical Problem	22/04/2022	22/04/2022
									A070101	Failure to Capture	22/04/2022	22/04/2022
									A0709	Device Sensing Problem	22/04/2022	22/04/2022

									A0712	Pacing Problem	22/04/2022	22/04/2022
									A072201	High Impedance	22/04/2022	22/04/2022
								F26	No Health Consequences or Impact	A05	Mechanical Problem	22/04/2022	22/04/2022
										A070101	Failure to Capture	22/04/2022	22/04/2022
										A0709	Device Sensing Problem	22/04/2022	22/04/2022
										A0712	Pacing Problem	22/04/2022	22/04/2022
										A072201	High Impedance	22/04/2022	22/04/2022
						CT2727	Active implant devices/ancillaries	E2403	No Clinical Signs, Symptoms or Conditions	A05	Mechanical Problem	22/04/2022	22/04/2022
										A070101	Failure to Capture	22/04/2022	22/04/2022
										A0709	Device Sensing Problem	22/04/2022	22/04/2022
										A0712	Pacing Problem	22/04/2022	22/04/2022
										A072201	High Impedance	22/04/2022	22/04/2022
								F26	No Health Consequences or Impact	A05	Mechanical Problem	22/04/2022	22/04/2022
										A070101	Failure to Capture	22/04/2022	22/04/2022
										A0709	Device Sensing Problem	22/04/2022	22/04/2022
										A0712	Pacing Problem	22/04/2022	22/04/2022
										A072201	High Impedance	22/04/2022	22/04/2022
						CT2750	Implantable cardiac electrical stimulators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A05	Mechanical Problem	22/04/2022	22/04/2022
										A070101	Failure to Capture	22/04/2022	22/04/2022
										A0709	Device Sensing Problem	22/04/2022	22/04/2022
										A0712	Pacing Problem	22/04/2022	22/04/2022
										A072201	High Impedance	22/04/2022	22/04/2022
								F26	No Health Consequences or Impact	A05	Mechanical Problem	22/04/2022	22/04/2022
										A070101	Failure to Capture	22/04/2022	22/04/2022
										A0709	Device Sensing Problem	22/04/2022	22/04/2022
										A0712	Pacing Problem	22/04/2022	22/04/2022
										A072201	High Impedance	22/04/2022	22/04/2022
						CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A05	Mechanical Problem	22/04/2022	22/04/2022
										A070101	Failure to Capture	22/04/2022	22/04/2022
										A0709	Device Sensing Problem	22/04/2022	22/04/2022
										A0712	Pacing Problem	22/04/2022	22/04/2022
										A072201	High Impedance	22/04/2022	22/04/2022
								F26	No Health Consequences or Impact	A05	Mechanical Problem	22/04/2022	22/04/2022
										A070101	Failure to Capture	22/04/2022	22/04/2022
										A0709	Device Sensing Problem	22/04/2022	22/04/2022
										A0712	Pacing Problem	22/04/2022	22/04/2022
										A072201	High Impedance	22/04/2022	22/04/2022
						CT2985	Active	E2403	No Clinical Signs, Symptoms or Conditions	A05	Mechanical Problem	22/04/2022	22/04/2022
										A070101	Failure to Capture	22/04/2022	22/04/2022
										A0709	Device Sensing Problem	22/04/2022	22/04/2022
										A0712	Pacing Problem	22/04/2022	22/04/2022
										A072201	High Impedance	22/04/2022	22/04/2022
								F26	No Health Consequences or Impact	A05	Mechanical Problem	22/04/2022	22/04/2022
										A070101	Failure to Capture	22/04/2022	22/04/2022
										A0709	Device Sensing Problem	22/04/2022	22/04/2022
										A0712	Pacing Problem	22/04/2022	22/04/2022
										A072201	High Impedance	22/04/2022	22/04/2022
						CT3001	Implantable metal	E2403	No Clinical Signs, Symptoms or Conditions	A05	Mechanical Problem	22/04/2022	22/04/2022
										A070101	Failure to Capture	22/04/2022	22/04/2022
										A0709	Device Sensing Problem	22/04/2022	22/04/2022
										A0712	Pacing Problem	22/04/2022	22/04/2022
										A072201	High Impedance	22/04/2022	22/04/2022
								F26	No Health Consequences or Impact	A05	Mechanical Problem	22/04/2022	22/04/2022
										A070101	Failure to Capture	22/04/2022	22/04/2022
										A0709	Device Sensing Problem	22/04/2022	22/04/2022
										A0712	Pacing Problem	22/04/2022	22/04/2022
										A072201	High Impedance	22/04/2022	22/04/2022
						CT3002	Implantable synthetic polymer	E2403	No Clinical Signs, Symptoms or Conditions	A05	Mechanical Problem	22/04/2022	22/04/2022
										A070101	Failure to Capture	22/04/2022	22/04/2022
										A0709	Device Sensing Problem	22/04/2022	22/04/2022
										A0712	Pacing Problem	22/04/2022	22/04/2022
										A072201	High Impedance	22/04/2022	22/04/2022
								F26	No Health Consequences or Impact	A05	Mechanical Problem	22/04/2022	22/04/2022
										A070101	Failure to Capture	22/04/2022	22/04/2022
										A0709	Device Sensing Problem	22/04/2022	22/04/2022
										A0712	Pacing Problem	22/04/2022	22/04/2022
										A072201	High Impedance	22/04/2022	22/04/2022
						CT3006	Implantable inorganic material	E2403	No Clinical Signs, Symptoms or Conditions	A05	Mechanical Problem	22/04/2022	22/04/2022
										A070101	Failure to Capture	22/04/2022	22/04/2022
										A0709	Device Sensing Problem	22/04/2022	22/04/2022
										A0712	Pacing Problem	22/04/2022	22/04/2022
										A072201	High Impedance	22/04/2022	22/04/2022
								F26	No Health Consequences or Impact	A05	Mechanical Problem	22/04/2022	22/04/2022
										A070101	Failure to Capture	22/04/2022	22/04/2022
										A0709	Device Sensing Problem	22/04/2022	22/04/2022
										A0712	Pacing Problem	22/04/2022	22/04/2022
										A072201	High Impedance	22/04/2022	22/04/2022
						CT314	Electrically powered	E2403	No Clinical Signs, Symptoms or Conditions	A05	Mechanical Problem	22/04/2022	22/04/2022
										A070101	Failure to Capture	22/04/2022	22/04/2022
										A0709	Device Sensing Problem	22/04/2022	22/04/2022
										A0712	Pacing Problem	22/04/2022	22/04/2022
										A072201	High Impedance	22/04/2022	22/04/2022
								F26	No Health Consequences or Impact	A05	Mechanical Problem	22/04/2022	22/04/2022
										A070101	Failure to Capture	22/04/2022	22/04/2022
										A0709	Device Sensing Problem	22/04/2022	22/04/2022
										A0712	Pacing Problem	22/04/2022	22/04/2022
										A072201	High Impedance	22/04/2022	22/04/2022
						CT321	Long-term surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A05	Mechanical Problem	22/04/2022	22/04/2022

									A070101	Failure to Capture	22/04/2022	22/04/2022
									A0709	Device Sensing Problem	22/04/2022	22/04/2022
									A0712	Pacing Problem	22/04/2022	22/04/2022
									A072201	High Impedance	22/04/2022	22/04/2022
							F26	No Health Consequences or Impact	A05	Mechanical Problem	22/04/2022	22/04/2022
									A070101	Failure to Capture	22/04/2022	22/04/2022
									A0709	Device Sensing Problem	22/04/2022	22/04/2022
									A0712	Pacing Problem	22/04/2022	22/04/2022
									A072201	High Impedance	22/04/2022	22/04/2022
					CT449	Generators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A05	Mechanical Problem	22/04/2022	22/04/2022
									A070101	Failure to Capture	22/04/2022	22/04/2022
									A0709	Device Sensing Problem	22/04/2022	22/04/2022
									A0712	Pacing Problem	22/04/2022	22/04/2022
									A072201	High Impedance	22/04/2022	22/04/2022
							F26	No Health Consequences or Impact	A05	Mechanical Problem	22/04/2022	22/04/2022
									A070101	Failure to Capture	22/04/2022	22/04/2022
									A0709	Device Sensing Problem	22/04/2022	22/04/2022
									A0712	Pacing Problem	22/04/2022	22/04/2022
									A072201	High Impedance	22/04/2022	22/04/2022
					CT981	Single use	E2403	No Clinical Signs, Symptoms or Conditions	A05	Mechanical Problem	22/04/2022	22/04/2022
									A070101	Failure to Capture	22/04/2022	22/04/2022
									A0709	Device Sensing Problem	22/04/2022	22/04/2022
									A0712	Pacing Problem	22/04/2022	22/04/2022
									A072201	High Impedance	22/04/2022	22/04/2022
							F26	No Health Consequences or Impact	A05	Mechanical Problem	22/04/2022	22/04/2022
									A070101	Failure to Capture	22/04/2022	22/04/2022
									A0709	Device Sensing Problem	22/04/2022	22/04/2022
									A0712	Pacing Problem	22/04/2022	22/04/2022
									A072201	High Impedance	22/04/2022	22/04/2022
					CT983	Surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A05	Mechanical Problem	22/04/2022	22/04/2022
									A070101	Failure to Capture	22/04/2022	22/04/2022
									A0709	Device Sensing Problem	22/04/2022	22/04/2022
									A0712	Pacing Problem	22/04/2022	22/04/2022
									A072201	High Impedance	22/04/2022	22/04/2022
							F26	No Health Consequences or Impact	A05	Mechanical Problem	22/04/2022	22/04/2022
									A070101	Failure to Capture	22/04/2022	22/04/2022
									A0709	Device Sensing Problem	22/04/2022	22/04/2022
									A0712	Pacing Problem	22/04/2022	22/04/2022
									A072201	High Impedance	22/04/2022	22/04/2022
6574	2022	All other reportable incidents	17882	Physiologic-monitoring defibrillation system	CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A0721	Circuit Failure		
							F27	Problem Identified During Non-Clinical Procedure	A0721	Circuit Failure		
6575	2022	All other reportable incidents	35223	Endocardial/interventricular septal pacing lead	CT1684	Cardiac pacemakers and associated devices	E0112	Dizziness	A0401	Break	11/05/2022	11/05/2022
									A070909	Oversensing	11/05/2022	11/05/2022
									A072201	High Impedance	11/05/2022	11/05/2022
									A090801	Signal Artifact/Noise	11/05/2022	11/05/2022
							E2403	No Clinical Signs, Symptoms or Conditions	A0401	Break	11/05/2022	11/05/2022
									A070909	Oversensing	11/05/2022	11/05/2022
									A072201	High Impedance	11/05/2022	11/05/2022
									A090801	Signal Artifact/Noise	11/05/2022	11/05/2022
							F08	Hospitalization or Prolonged Hospitalization	A0401	Break	11/05/2022	11/05/2022
									A070909	Oversensing	11/05/2022	11/05/2022
									A072201	High Impedance	11/05/2022	11/05/2022
									A090801	Signal Artifact/Noise	11/05/2022	11/05/2022
							F12	Serious Injury/ Illness/ Impairment	A0401	Break	11/05/2022	11/05/2022
									A070909	Oversensing	11/05/2022	11/05/2022
									A072201	High Impedance	11/05/2022	11/05/2022
									A090801	Signal Artifact/Noise	11/05/2022	11/05/2022
							F19	Surgical Intervention	A0401	Break	11/05/2022	11/05/2022
									A070909	Oversensing	11/05/2022	11/05/2022
									A072201	High Impedance	11/05/2022	11/05/2022
									A090801	Signal Artifact/Noise	11/05/2022	11/05/2022
							F23	Unexpected Medical Intervention	A0401	Break	11/05/2022	11/05/2022
									A070909	Oversensing	11/05/2022	11/05/2022
									A072201	High Impedance	11/05/2022	11/05/2022
									A090801	Signal Artifact/Noise	11/05/2022	11/05/2022
6576	2022	All other reportable incidents	35852	Single-chamber implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A070910	Undersensing		
									A072201	High Impedance		
							F22	Unexpected Diagnostic Intervention	A070910	Undersensing		
									A072201	High Impedance		
6577	2021	All other reportable incidents	35223	Endocardial/interventricular septal pacing lead	CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A0401	Break		
									A070910	Undersensing		
									A0722	Impedance Problem		
							F26	No Health Consequences or Impact	A0401	Break		
									A070910	Undersensing		
									A0722	Impedance Problem		
6578	2022	All other reportable incidents	36241	Pacemaker/defibrillator lead	CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070910	Undersensing		
									A072201	High Impedance		
							F22	Unexpected Diagnostic Intervention	A070910	Undersensing		
									A072201	High Impedance		
6579	2022	All other reportable incidents	47265	Dual-chamber implantable pacemaker, rate-responsive	CT141	Electrodes/Leads and associated devices	E011903	Syncope/Fainting	A070909	Oversensing	04/05/2022	04/05/2022
									A0712	Pacing Problem	04/05/2022	04/05/2022
									A0908	Incorrect, Inadequate or Imprecise Result or Readings	04/05/2022	04/05/2022
							E060101	Asystole	A070909	Oversensing	04/05/2022	04/05/2022
									A0712	Pacing Problem	04/05/2022	04/05/2022
									A0908	Incorrect, Inadequate or Imprecise Result or Readings	04/05/2022	04/05/2022
							E2403	No Clinical Signs, Symptoms or Conditions	A070909	Oversensing	04/05/2022	04/05/2022

								A0712	Pacing Problem	04/05/2022	04/05/2022
								A0908	Incorrect, Inadequate or Imprecise Result or Readings	04/05/2022	04/05/2022
						F08	Hospitalization or Prolonged Hospitalization	A070909	Oversensing	04/05/2022	04/05/2022
								A0712	Pacing Problem	04/05/2022	04/05/2022
								A0908	Incorrect, Inadequate or Imprecise Result or Readings	04/05/2022	04/05/2022
						F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A070909	Oversensing	04/05/2022	04/05/2022
								A0712	Pacing Problem	04/05/2022	04/05/2022
								A0908	Incorrect, Inadequate or Imprecise Result or Readings	04/05/2022	04/05/2022
						F1905	Device Revision or Replacement	A070909	Oversensing	04/05/2022	04/05/2022
								A0712	Pacing Problem	04/05/2022	04/05/2022
								A0908	Incorrect, Inadequate or Imprecise Result or Readings	04/05/2022	04/05/2022
					CT145	Cardiovascular devices	E011903	Syncope/Fainting	A070909	Oversensing	04/05/2022	04/05/2022
									A0712	Pacing Problem	04/05/2022	04/05/2022
									A0908	Incorrect, Inadequate or Imprecise Result or Readings	04/05/2022	04/05/2022
							E060101	Asystole	A070909	Oversensing	04/05/2022	04/05/2022
									A0712	Pacing Problem	04/05/2022	04/05/2022
									A0908	Incorrect, Inadequate or Imprecise Result or Readings	04/05/2022	04/05/2022
							E2403	No Clinical Signs, Symptoms or Conditions	A070909	Oversensing	04/05/2022	04/05/2022
									A0712	Pacing Problem	04/05/2022	04/05/2022
									A0908	Incorrect, Inadequate or Imprecise Result or Readings	04/05/2022	04/05/2022
							F08	Hospitalization or Prolonged Hospitalization	A070909	Oversensing	04/05/2022	04/05/2022
									A0712	Pacing Problem	04/05/2022	04/05/2022
									A0908	Incorrect, Inadequate or Imprecise Result or Readings	04/05/2022	04/05/2022
							F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A070909	Oversensing	04/05/2022	04/05/2022
									A0712	Pacing Problem	04/05/2022	04/05/2022
									A0908	Incorrect, Inadequate or Imprecise Result or Readings	04/05/2022	04/05/2022
							F1905	Device Revision or Replacement	A070909	Oversensing	04/05/2022	04/05/2022
									A0712	Pacing Problem	04/05/2022	04/05/2022
									A0908	Incorrect, Inadequate or Imprecise Result or Readings	04/05/2022	04/05/2022
					CT1684	Cardiac pacemakers and associated devices	E011903	Syncope/Fainting	A070909	Oversensing	04/05/2022	04/05/2022
									A0712	Pacing Problem	04/05/2022	04/05/2022
									A0908	Incorrect, Inadequate or Imprecise Result or Readings	04/05/2022	04/05/2022
							E060101	Asystole	A070909	Oversensing	04/05/2022	04/05/2022
									A0712	Pacing Problem	04/05/2022	04/05/2022
									A0908	Incorrect, Inadequate or Imprecise Result or Readings	04/05/2022	04/05/2022
							E2403	No Clinical Signs, Symptoms or Conditions	A070909	Oversensing	04/05/2022	04/05/2022
									A0712	Pacing Problem	04/05/2022	04/05/2022
									A0908	Incorrect, Inadequate or Imprecise Result or Readings	04/05/2022	04/05/2022
							F08	Hospitalization or Prolonged Hospitalization	A070909	Oversensing	04/05/2022	04/05/2022
									A0712	Pacing Problem	04/05/2022	04/05/2022
									A0908	Incorrect, Inadequate or Imprecise Result or Readings	04/05/2022	04/05/2022
							F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A070909	Oversensing	04/05/2022	04/05/2022
									A0712	Pacing Problem	04/05/2022	04/05/2022
									A0908	Incorrect, Inadequate or Imprecise Result or Readings	04/05/2022	04/05/2022
							F1905	Device Revision or Replacement	A070909	Oversensing	04/05/2022	04/05/2022
									A0712	Pacing Problem	04/05/2022	04/05/2022
									A0908	Incorrect, Inadequate or Imprecise Result or Readings	04/05/2022	04/05/2022
					CT1685	Cardiac pacemakers	E011903	Syncope/Fainting	A070909	Oversensing	04/05/2022	04/05/2022
									A0712	Pacing Problem	04/05/2022	04/05/2022
									A0908	Incorrect, Inadequate or Imprecise Result or Readings	04/05/2022	04/05/2022
							E060101	Asystole	A070909	Oversensing	04/05/2022	04/05/2022
									A0712	Pacing Problem	04/05/2022	04/05/2022
									A0908	Incorrect, Inadequate or Imprecise Result or Readings	04/05/2022	04/05/2022
							E2403	No Clinical Signs, Symptoms or Conditions	A070909	Oversensing	04/05/2022	04/05/2022
									A0712	Pacing Problem	04/05/2022	04/05/2022
									A0908	Incorrect, Inadequate or Imprecise Result or Readings	04/05/2022	04/05/2022
							F08	Hospitalization or Prolonged Hospitalization	A070909	Oversensing	04/05/2022	04/05/2022
									A0712	Pacing Problem	04/05/2022	04/05/2022
									A0908	Incorrect, Inadequate or Imprecise Result or Readings	04/05/2022	04/05/2022
							F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A070909	Oversensing	04/05/2022	04/05/2022
									A0712	Pacing Problem	04/05/2022	04/05/2022
									A0908	Incorrect, Inadequate or Imprecise Result or Readings	04/05/2022	04/05/2022
							F1905	Device Revision or Replacement	A070909	Oversensing	04/05/2022	04/05/2022
									A0712	Pacing Problem	04/05/2022	04/05/2022
									A0908	Incorrect, Inadequate or Imprecise Result or Readings	04/05/2022	04/05/2022
					CT233	Surgical	E011903	Syncope/Fainting	A070909	Oversensing	04/05/2022	04/05/2022
									A0712	Pacing Problem	04/05/2022	04/05/2022
									A0908	Incorrect, Inadequate or Imprecise Result or Readings	04/05/2022	04/05/2022
							E060101	Asystole	A070909	Oversensing	04/05/2022	04/05/2022
									A0712	Pacing Problem	04/05/2022	04/05/2022

								A0908	Incorrect, Inadequate or Imprecise Result or Readings	04/05/2022	04/05/2022
							E2403	No Clinical Signs, Symptoms or Conditions	A070909	Oversensing	04/05/2022	04/05/2022
									A0712	Pacing Problem	04/05/2022	04/05/2022
									A0908	Incorrect, Inadequate or Imprecise Result or Readings	04/05/2022	04/05/2022
							F08	Hospitalization or Prolonged Hospitalization	A070909	Oversensing	04/05/2022	04/05/2022
									A0712	Pacing Problem	04/05/2022	04/05/2022
									A0908	Incorrect, Inadequate or Imprecise Result or Readings	04/05/2022	04/05/2022
							F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A070909	Oversensing	04/05/2022	04/05/2022
									A0712	Pacing Problem	04/05/2022	04/05/2022
									A0908	Incorrect, Inadequate or Imprecise Result or Readings	04/05/2022	04/05/2022
							F1905	Device Revision or Replacement	A070909	Oversensing	04/05/2022	04/05/2022
									A0712	Pacing Problem	04/05/2022	04/05/2022
									A0908	Incorrect, Inadequate or Imprecise Result or Readings	04/05/2022	04/05/2022
					CT2406	Implantable	E011903	Syncope/Fainting	A070909	Oversensing	04/05/2022	04/05/2022
									A0712	Pacing Problem	04/05/2022	04/05/2022
									A0908	Incorrect, Inadequate or Imprecise Result or Readings	04/05/2022	04/05/2022
							E060101	Asystole	A070909	Oversensing	04/05/2022	04/05/2022
									A0712	Pacing Problem	04/05/2022	04/05/2022
									A0908	Incorrect, Inadequate or Imprecise Result or Readings	04/05/2022	04/05/2022
							E2403	No Clinical Signs, Symptoms or Conditions	A070909	Oversensing	04/05/2022	04/05/2022
									A0712	Pacing Problem	04/05/2022	04/05/2022
									A0908	Incorrect, Inadequate or Imprecise Result or Readings	04/05/2022	04/05/2022
							F08	Hospitalization or Prolonged Hospitalization	A070909	Oversensing	04/05/2022	04/05/2022
									A0712	Pacing Problem	04/05/2022	04/05/2022
									A0908	Incorrect, Inadequate or Imprecise Result or Readings	04/05/2022	04/05/2022
							F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A070909	Oversensing	04/05/2022	04/05/2022
									A0712	Pacing Problem	04/05/2022	04/05/2022
									A0908	Incorrect, Inadequate or Imprecise Result or Readings	04/05/2022	04/05/2022
							F1905	Device Revision or Replacement	A070909	Oversensing	04/05/2022	04/05/2022
									A0712	Pacing Problem	04/05/2022	04/05/2022
									A0908	Incorrect, Inadequate or Imprecise Result or Readings	04/05/2022	04/05/2022
					CT2407	Active implantable	E011903	Syncope/Fainting	A070909	Oversensing	04/05/2022	04/05/2022
									A0712	Pacing Problem	04/05/2022	04/05/2022
									A0908	Incorrect, Inadequate or Imprecise Result or Readings	04/05/2022	04/05/2022
							E060101	Asystole	A070909	Oversensing	04/05/2022	04/05/2022
									A0712	Pacing Problem	04/05/2022	04/05/2022
									A0908	Incorrect, Inadequate or Imprecise Result or Readings	04/05/2022	04/05/2022
							E2403	No Clinical Signs, Symptoms or Conditions	A070909	Oversensing	04/05/2022	04/05/2022
									A0712	Pacing Problem	04/05/2022	04/05/2022
									A0908	Incorrect, Inadequate or Imprecise Result or Readings	04/05/2022	04/05/2022
							F08	Hospitalization or Prolonged Hospitalization	A070909	Oversensing	04/05/2022	04/05/2022
									A0712	Pacing Problem	04/05/2022	04/05/2022
									A0908	Incorrect, Inadequate or Imprecise Result or Readings	04/05/2022	04/05/2022
							F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A070909	Oversensing	04/05/2022	04/05/2022
									A0712	Pacing Problem	04/05/2022	04/05/2022
									A0908	Incorrect, Inadequate or Imprecise Result or Readings	04/05/2022	04/05/2022
							F1905	Device Revision or Replacement	A070909	Oversensing	04/05/2022	04/05/2022
									A0712	Pacing Problem	04/05/2022	04/05/2022
									A0908	Incorrect, Inadequate or Imprecise Result or Readings	04/05/2022	04/05/2022
					CT257	Implantable cardiac pacemakers	E011903	Syncope/Fainting	A070909	Oversensing	04/05/2022	04/05/2022
									A0712	Pacing Problem	04/05/2022	04/05/2022
									A0908	Incorrect, Inadequate or Imprecise Result or Readings	04/05/2022	04/05/2022
							E060101	Asystole	A070909	Oversensing	04/05/2022	04/05/2022
									A0712	Pacing Problem	04/05/2022	04/05/2022
									A0908	Incorrect, Inadequate or Imprecise Result or Readings	04/05/2022	04/05/2022
							E2403	No Clinical Signs, Symptoms or Conditions	A070909	Oversensing	04/05/2022	04/05/2022
									A0712	Pacing Problem	04/05/2022	04/05/2022
									A0908	Incorrect, Inadequate or Imprecise Result or Readings	04/05/2022	04/05/2022
							F08	Hospitalization or Prolonged Hospitalization	A070909	Oversensing	04/05/2022	04/05/2022
									A0712	Pacing Problem	04/05/2022	04/05/2022
									A0908	Incorrect, Inadequate or Imprecise Result or Readings	04/05/2022	04/05/2022
							F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A070909	Oversensing	04/05/2022	04/05/2022
									A0712	Pacing Problem	04/05/2022	04/05/2022
									A0908	Incorrect, Inadequate or Imprecise Result or Readings	04/05/2022	04/05/2022
							F1905	Device Revision or Replacement	A070909	Oversensing	04/05/2022	04/05/2022
									A0712	Pacing Problem	04/05/2022	04/05/2022
									A0908	Incorrect, Inadequate or Imprecise Result or Readings	04/05/2022	04/05/2022
					CT2727	Active implant devices/ancillaries	E011903	Syncope/Fainting	A070909	Oversensing	04/05/2022	04/05/2022
									A0712	Pacing Problem	04/05/2022	04/05/2022

								A0908	Incorrect, Inadequate or Imprecise Result or Readings	04/05/2022	04/05/2022
							E060101	Asystole	A070909	Oversensing	04/05/2022	04/05/2022
									A0712	Pacing Problem	04/05/2022	04/05/2022
									A0908	Incorrect, Inadequate or Imprecise Result or Readings	04/05/2022	04/05/2022
							E2403	No Clinical Signs, Symptoms or Conditions	A070909	Oversensing	04/05/2022	04/05/2022
									A0712	Pacing Problem	04/05/2022	04/05/2022
									A0908	Incorrect, Inadequate or Imprecise Result or Readings	04/05/2022	04/05/2022
							F08	Hospitalization or Prolonged Hospitalization	A070909	Oversensing	04/05/2022	04/05/2022
									A0712	Pacing Problem	04/05/2022	04/05/2022
									A0908	Incorrect, Inadequate or Imprecise Result or Readings	04/05/2022	04/05/2022
							F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A070909	Oversensing	04/05/2022	04/05/2022
									A0712	Pacing Problem	04/05/2022	04/05/2022
									A0908	Incorrect, Inadequate or Imprecise Result or Readings	04/05/2022	04/05/2022
							F1905	Device Revision or Replacement	A070909	Oversensing	04/05/2022	04/05/2022
									A0712	Pacing Problem	04/05/2022	04/05/2022
									A0908	Incorrect, Inadequate or Imprecise Result or Readings	04/05/2022	04/05/2022
						CT2750	Implantable cardiac electrical stimulators and associated devices	E011903	Syncope/Fainting	A070909	Oversensing	04/05/2022	04/05/2022
										A0712	Pacing Problem	04/05/2022	04/05/2022
										A0908	Incorrect, Inadequate or Imprecise Result or Readings	04/05/2022	04/05/2022
								E060101	Asystole	A070909	Oversensing	04/05/2022	04/05/2022
										A0712	Pacing Problem	04/05/2022	04/05/2022
										A0908	Incorrect, Inadequate or Imprecise Result or Readings	04/05/2022	04/05/2022
								E2403	No Clinical Signs, Symptoms or Conditions	A070909	Oversensing	04/05/2022	04/05/2022
										A0712	Pacing Problem	04/05/2022	04/05/2022
										A0908	Incorrect, Inadequate or Imprecise Result or Readings	04/05/2022	04/05/2022
								F08	Hospitalization or Prolonged Hospitalization	A070909	Oversensing	04/05/2022	04/05/2022
										A0712	Pacing Problem	04/05/2022	04/05/2022
										A0908	Incorrect, Inadequate or Imprecise Result or Readings	04/05/2022	04/05/2022
								F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A070909	Oversensing	04/05/2022	04/05/2022
										A0712	Pacing Problem	04/05/2022	04/05/2022
										A0908	Incorrect, Inadequate or Imprecise Result or Readings	04/05/2022	04/05/2022
								F1905	Device Revision or Replacement	A070909	Oversensing	04/05/2022	04/05/2022
										A0712	Pacing Problem	04/05/2022	04/05/2022
										A0908	Incorrect, Inadequate or Imprecise Result or Readings	04/05/2022	04/05/2022
						CT2866	Implantable cardiac electrical stimulators	E011903	Syncope/Fainting	A070909	Oversensing	04/05/2022	04/05/2022
										A0712	Pacing Problem	04/05/2022	04/05/2022
										A0908	Incorrect, Inadequate or Imprecise Result or Readings	04/05/2022	04/05/2022
								E060101	Asystole	A070909	Oversensing	04/05/2022	04/05/2022
										A0712	Pacing Problem	04/05/2022	04/05/2022
										A0908	Incorrect, Inadequate or Imprecise Result or Readings	04/05/2022	04/05/2022
								E2403	No Clinical Signs, Symptoms or Conditions	A070909	Oversensing	04/05/2022	04/05/2022
										A0712	Pacing Problem	04/05/2022	04/05/2022
										A0908	Incorrect, Inadequate or Imprecise Result or Readings	04/05/2022	04/05/2022
								F08	Hospitalization or Prolonged Hospitalization	A070909	Oversensing	04/05/2022	04/05/2022
										A0712	Pacing Problem	04/05/2022	04/05/2022
										A0908	Incorrect, Inadequate or Imprecise Result or Readings	04/05/2022	04/05/2022
								F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A070909	Oversensing	04/05/2022	04/05/2022
										A0712	Pacing Problem	04/05/2022	04/05/2022
										A0908	Incorrect, Inadequate or Imprecise Result or Readings	04/05/2022	04/05/2022
								F1905	Device Revision or Replacement	A070909	Oversensing	04/05/2022	04/05/2022
										A0712	Pacing Problem	04/05/2022	04/05/2022
										A0908	Incorrect, Inadequate or Imprecise Result or Readings	04/05/2022	04/05/2022
						CT2985	Active	E011903	Syncope/Fainting	A070909	Oversensing	04/05/2022	04/05/2022
										A0712	Pacing Problem	04/05/2022	04/05/2022
										A0908	Incorrect, Inadequate or Imprecise Result or Readings	04/05/2022	04/05/2022
								E060101	Asystole	A070909	Oversensing	04/05/2022	04/05/2022
										A0712	Pacing Problem	04/05/2022	04/05/2022
										A0908	Incorrect, Inadequate or Imprecise Result or Readings	04/05/2022	04/05/2022
								E2403	No Clinical Signs, Symptoms or Conditions	A070909	Oversensing	04/05/2022	04/05/2022
										A0712	Pacing Problem	04/05/2022	04/05/2022
										A0908	Incorrect, Inadequate or Imprecise Result or Readings	04/05/2022	04/05/2022
								F08	Hospitalization or Prolonged Hospitalization	A070909	Oversensing	04/05/2022	04/05/2022
										A0712	Pacing Problem	04/05/2022	04/05/2022
										A0908	Incorrect, Inadequate or Imprecise Result or Readings	04/05/2022	04/05/2022
								F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A070909	Oversensing	04/05/2022	04/05/2022
										A0712	Pacing Problem	04/05/2022	04/05/2022
										A0908	Incorrect, Inadequate or Imprecise Result or Readings	04/05/2022	04/05/2022
								F1905	Device Revision or Replacement	A070909	Oversensing	04/05/2022	04/05/2022
										A0712	Pacing Problem	04/05/2022	04/05/2022

								A0908	Incorrect, Inadequate or Imprecise Result or Readings	04/05/2022	04/05/2022
					CT3001	Implantable metal	E011903	Syncope/Fainting	A070909	Oversensing	04/05/2022	04/05/2022
									A0712	Pacing Problem	04/05/2022	04/05/2022
									A0908	Incorrect, Inadequate or Imprecise Result or Readings	04/05/2022	04/05/2022
							E060101	Asystole	A070909	Oversensing	04/05/2022	04/05/2022
									A0712	Pacing Problem	04/05/2022	04/05/2022
									A0908	Incorrect, Inadequate or Imprecise Result or Readings	04/05/2022	04/05/2022
							E2403	No Clinical Signs, Symptoms or Conditions	A070909	Oversensing	04/05/2022	04/05/2022
									A0712	Pacing Problem	04/05/2022	04/05/2022
									A0908	Incorrect, Inadequate or Imprecise Result or Readings	04/05/2022	04/05/2022
							F08	Hospitalization or Prolonged Hospitalization	A070909	Oversensing	04/05/2022	04/05/2022
									A0712	Pacing Problem	04/05/2022	04/05/2022
									A0908	Incorrect, Inadequate or Imprecise Result or Readings	04/05/2022	04/05/2022
							F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A070909	Oversensing	04/05/2022	04/05/2022
									A0712	Pacing Problem	04/05/2022	04/05/2022
									A0908	Incorrect, Inadequate or Imprecise Result or Readings	04/05/2022	04/05/2022
							F1905	Device Revision or Replacement	A070909	Oversensing	04/05/2022	04/05/2022
									A0712	Pacing Problem	04/05/2022	04/05/2022
									A0908	Incorrect, Inadequate or Imprecise Result or Readings	04/05/2022	04/05/2022
					CT3002	Implantable synthetic polymer	E011903	Syncope/Fainting	A070909	Oversensing	04/05/2022	04/05/2022
									A0712	Pacing Problem	04/05/2022	04/05/2022
									A0908	Incorrect, Inadequate or Imprecise Result or Readings	04/05/2022	04/05/2022
							E060101	Asystole	A070909	Oversensing	04/05/2022	04/05/2022
									A0712	Pacing Problem	04/05/2022	04/05/2022
									A0908	Incorrect, Inadequate or Imprecise Result or Readings	04/05/2022	04/05/2022
							E2403	No Clinical Signs, Symptoms or Conditions	A070909	Oversensing	04/05/2022	04/05/2022
									A0712	Pacing Problem	04/05/2022	04/05/2022
									A0908	Incorrect, Inadequate or Imprecise Result or Readings	04/05/2022	04/05/2022
							F08	Hospitalization or Prolonged Hospitalization	A070909	Oversensing	04/05/2022	04/05/2022
									A0712	Pacing Problem	04/05/2022	04/05/2022
									A0908	Incorrect, Inadequate or Imprecise Result or Readings	04/05/2022	04/05/2022
							F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A070909	Oversensing	04/05/2022	04/05/2022
									A0712	Pacing Problem	04/05/2022	04/05/2022
									A0908	Incorrect, Inadequate or Imprecise Result or Readings	04/05/2022	04/05/2022
							F1905	Device Revision or Replacement	A070909	Oversensing	04/05/2022	04/05/2022
									A0712	Pacing Problem	04/05/2022	04/05/2022
									A0908	Incorrect, Inadequate or Imprecise Result or Readings	04/05/2022	04/05/2022
					CT3006	Implantable inorganic material	E011903	Syncope/Fainting	A070909	Oversensing	04/05/2022	04/05/2022
									A0712	Pacing Problem	04/05/2022	04/05/2022
									A0908	Incorrect, Inadequate or Imprecise Result or Readings	04/05/2022	04/05/2022
							E060101	Asystole	A070909	Oversensing	04/05/2022	04/05/2022
									A0712	Pacing Problem	04/05/2022	04/05/2022
									A0908	Incorrect, Inadequate or Imprecise Result or Readings	04/05/2022	04/05/2022
							E2403	No Clinical Signs, Symptoms or Conditions	A070909	Oversensing	04/05/2022	04/05/2022
									A0712	Pacing Problem	04/05/2022	04/05/2022
									A0908	Incorrect, Inadequate or Imprecise Result or Readings	04/05/2022	04/05/2022
							F08	Hospitalization or Prolonged Hospitalization	A070909	Oversensing	04/05/2022	04/05/2022
									A0712	Pacing Problem	04/05/2022	04/05/2022
									A0908	Incorrect, Inadequate or Imprecise Result or Readings	04/05/2022	04/05/2022
							F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A070909	Oversensing	04/05/2022	04/05/2022
									A0712	Pacing Problem	04/05/2022	04/05/2022
									A0908	Incorrect, Inadequate or Imprecise Result or Readings	04/05/2022	04/05/2022
							F1905	Device Revision or Replacement	A070909	Oversensing	04/05/2022	04/05/2022
									A0712	Pacing Problem	04/05/2022	04/05/2022
									A0908	Incorrect, Inadequate or Imprecise Result or Readings	04/05/2022	04/05/2022
					CT314	Electrically powered	E011903	Syncope/Fainting	A070909	Oversensing	04/05/2022	04/05/2022
									A0712	Pacing Problem	04/05/2022	04/05/2022
									A0908	Incorrect, Inadequate or Imprecise Result or Readings	04/05/2022	04/05/2022
							E060101	Asystole	A070909	Oversensing	04/05/2022	04/05/2022
									A0712	Pacing Problem	04/05/2022	04/05/2022
									A0908	Incorrect, Inadequate or Imprecise Result or Readings	04/05/2022	04/05/2022
							E2403	No Clinical Signs, Symptoms or Conditions	A070909	Oversensing	04/05/2022	04/05/2022
									A0712	Pacing Problem	04/05/2022	04/05/2022
									A0908	Incorrect, Inadequate or Imprecise Result or Readings	04/05/2022	04/05/2022
							F08	Hospitalization or Prolonged Hospitalization	A070909	Oversensing	04/05/2022	04/05/2022
									A0712	Pacing Problem	04/05/2022	04/05/2022
									A0908	Incorrect, Inadequate or Imprecise Result or Readings	04/05/2022	04/05/2022
							F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A070909	Oversensing	04/05/2022	04/05/2022
									A0712	Pacing Problem	04/05/2022	04/05/2022

								A0908	Incorrect, Inadequate or Imprecise Result or Readings	04/05/2022	04/05/2022
						F1905	Device Revision or Replacement	A070909	Oversensing	04/05/2022	04/05/2022
								A0712	Pacing Problem	04/05/2022	04/05/2022
								A0908	Incorrect, Inadequate or Imprecise Result or Readings	04/05/2022	04/05/2022
					CT321	Long-term surgical invasive	E011903	Syncope/Fainting	A070909	Oversensing	04/05/2022	04/05/2022
									A0712	Pacing Problem	04/05/2022	04/05/2022
									A0908	Incorrect, Inadequate or Imprecise Result or Readings	04/05/2022	04/05/2022
							E060101	Asystole	A070909	Oversensing	04/05/2022	04/05/2022
									A0712	Pacing Problem	04/05/2022	04/05/2022
									A0908	Incorrect, Inadequate or Imprecise Result or Readings	04/05/2022	04/05/2022
							E2403	No Clinical Signs, Symptoms or Conditions	A070909	Oversensing	04/05/2022	04/05/2022
									A0712	Pacing Problem	04/05/2022	04/05/2022
									A0908	Incorrect, Inadequate or Imprecise Result or Readings	04/05/2022	04/05/2022
							F08	Hospitalization or Prolonged Hospitalization	A070909	Oversensing	04/05/2022	04/05/2022
									A0712	Pacing Problem	04/05/2022	04/05/2022
									A0908	Incorrect, Inadequate or Imprecise Result or Readings	04/05/2022	04/05/2022
							F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A070909	Oversensing	04/05/2022	04/05/2022
									A0712	Pacing Problem	04/05/2022	04/05/2022
									A0908	Incorrect, Inadequate or Imprecise Result or Readings	04/05/2022	04/05/2022
							F1905	Device Revision or Replacement	A070909	Oversensing	04/05/2022	04/05/2022
									A0712	Pacing Problem	04/05/2022	04/05/2022
									A0908	Incorrect, Inadequate or Imprecise Result or Readings	04/05/2022	04/05/2022
					CT449	Generators and associated devices	E011903	Syncope/Fainting	A070909	Oversensing	04/05/2022	04/05/2022
									A0712	Pacing Problem	04/05/2022	04/05/2022
									A0908	Incorrect, Inadequate or Imprecise Result or Readings	04/05/2022	04/05/2022
							E060101	Asystole	A070909	Oversensing	04/05/2022	04/05/2022
									A0712	Pacing Problem	04/05/2022	04/05/2022
									A0908	Incorrect, Inadequate or Imprecise Result or Readings	04/05/2022	04/05/2022
							E2403	No Clinical Signs, Symptoms or Conditions	A070909	Oversensing	04/05/2022	04/05/2022
									A0712	Pacing Problem	04/05/2022	04/05/2022
									A0908	Incorrect, Inadequate or Imprecise Result or Readings	04/05/2022	04/05/2022
							F08	Hospitalization or Prolonged Hospitalization	A070909	Oversensing	04/05/2022	04/05/2022
									A0712	Pacing Problem	04/05/2022	04/05/2022
									A0908	Incorrect, Inadequate or Imprecise Result or Readings	04/05/2022	04/05/2022
							F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A070909	Oversensing	04/05/2022	04/05/2022
									A0712	Pacing Problem	04/05/2022	04/05/2022
									A0908	Incorrect, Inadequate or Imprecise Result or Readings	04/05/2022	04/05/2022
							F1905	Device Revision or Replacement	A070909	Oversensing	04/05/2022	04/05/2022
									A0712	Pacing Problem	04/05/2022	04/05/2022
									A0908	Incorrect, Inadequate or Imprecise Result or Readings	04/05/2022	04/05/2022
					CT981	Single use	E011903	Syncope/Fainting	A070909	Oversensing	04/05/2022	04/05/2022
									A0712	Pacing Problem	04/05/2022	04/05/2022
									A0908	Incorrect, Inadequate or Imprecise Result or Readings	04/05/2022	04/05/2022
							E060101	Asystole	A070909	Oversensing	04/05/2022	04/05/2022
									A0712	Pacing Problem	04/05/2022	04/05/2022
									A0908	Incorrect, Inadequate or Imprecise Result or Readings	04/05/2022	04/05/2022
							E2403	No Clinical Signs, Symptoms or Conditions	A070909	Oversensing	04/05/2022	04/05/2022
									A0712	Pacing Problem	04/05/2022	04/05/2022
									A0908	Incorrect, Inadequate or Imprecise Result or Readings	04/05/2022	04/05/2022
							F08	Hospitalization or Prolonged Hospitalization	A070909	Oversensing	04/05/2022	04/05/2022
									A0712	Pacing Problem	04/05/2022	04/05/2022
									A0908	Incorrect, Inadequate or Imprecise Result or Readings	04/05/2022	04/05/2022
							F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A070909	Oversensing	04/05/2022	04/05/2022
									A0712	Pacing Problem	04/05/2022	04/05/2022
									A0908	Incorrect, Inadequate or Imprecise Result or Readings	04/05/2022	04/05/2022
							F1905	Device Revision or Replacement	A070909	Oversensing	04/05/2022	04/05/2022
									A0712	Pacing Problem	04/05/2022	04/05/2022
									A0908	Incorrect, Inadequate or Imprecise Result or Readings	04/05/2022	04/05/2022
					CT983	Surgical invasive	E011903	Syncope/Fainting	A070909	Oversensing	04/05/2022	04/05/2022
									A0712	Pacing Problem	04/05/2022	04/05/2022
									A0908	Incorrect, Inadequate or Imprecise Result or Readings	04/05/2022	04/05/2022
							E060101	Asystole	A070909	Oversensing	04/05/2022	04/05/2022
									A0712	Pacing Problem	04/05/2022	04/05/2022
									A0908	Incorrect, Inadequate or Imprecise Result or Readings	04/05/2022	04/05/2022
							E2403	No Clinical Signs, Symptoms or Conditions	A070909	Oversensing	04/05/2022	04/05/2022
									A0712	Pacing Problem	04/05/2022	04/05/2022
									A0908	Incorrect, Inadequate or Imprecise Result or Readings	04/05/2022	04/05/2022
							F08	Hospitalization or Prolonged Hospitalization	A070909	Oversensing	04/05/2022	04/05/2022
									A0712	Pacing Problem	04/05/2022	04/05/2022

									A0908	Incorrect, Inadequate or Imprecise Result or Readings	04/05/2022	04/05/2022
							F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A070909	Oversensing	04/05/2022	04/05/2022
									A0712	Pacing Problem	04/05/2022	04/05/2022
									A0908	Incorrect, Inadequate or Imprecise Result or Readings	04/05/2022	04/05/2022
							F1905	Device Revision or Replacement	A070909	Oversensing	04/05/2022	04/05/2022
									A0712	Pacing Problem	04/05/2022	04/05/2022
									A0908	Incorrect, Inadequate or Imprecise Result or Readings	04/05/2022	04/05/2022
6580	2021	All other reportable incidents	17882	Physiologic-monitoring defibrillation system	CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A07	Electrical /Electronic Property Problem		
							F27	Problem Identified During Non-Clinical Procedure	A07	Electrical /Electronic Property Problem		
6581	2021	All other reportable incidents	47270	Cardiac resynchronization therapy implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	13/09/2021	13/09/2021
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery	13/09/2021	13/09/2021
6582	2021	All other reportable incidents	47267	Single-chamber implantable pacemaker, rate-responsive	CT141	Electrodes/Leads and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
							F24	Insufficient Health Impact Information	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
					CT145	Cardiovascular devices	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
							F24	Insufficient Health Impact Information	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
					CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
							F24	Insufficient Health Impact Information	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
					CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
							F24	Insufficient Health Impact Information	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
					CT233	Surgical	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
							F24	Insufficient Health Impact Information	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
					CT2406	Implantable	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
							F24	Insufficient Health Impact Information	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
					CT2407	Active implantable	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
							F24	Insufficient Health Impact Information	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
							F24	Insufficient Health Impact Information	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
					CT2727	Active implant devices/ancillaries	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
							F24	Insufficient Health Impact Information	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
					CT2750	Implantable cardiac electrical stimulators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
							F24	Insufficient Health Impact Information	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
							F24	Insufficient Health Impact Information	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
					CT2985	Active	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
							F24	Insufficient Health Impact Information	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
					CT3001	Implantable metal	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
							F24	Insufficient Health Impact Information	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
					CT3002	Implantable synthetic polymer	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
							F24	Insufficient Health Impact Information	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
					CT3006	Implantable inorganic material	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
							F24	Insufficient Health Impact Information	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
					CT314	Electrically powered	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
							F24	Insufficient Health Impact Information	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
					CT321	Long-term surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
							F24	Insufficient Health Impact Information	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
					CT449	Generators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
							F24	Insufficient Health Impact Information	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
					CT981	Single use	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		

							F24	Insufficient Health Impact Information	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
					CT983	Surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
							F24	Insufficient Health Impact Information	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
6583	2021	All other reportable incidents	35852	Single-chamber implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	11/11/2021	11/11/2021
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	11/11/2021	11/11/2021
6584	2021	All other reportable incidents	37265	Dual-chamber implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	09/09/2021	09/09/2021
							F1901	Additional Surgery	A070504	Premature Discharge of Battery	09/09/2021	09/09/2021
6585	2022	All other reportable incidents	35223	Endocardial/interventricular septal pacing lead	CT1684	Cardiac pacemakers and associated devices	E2328	Obstruction/Occlusion	A040101	Fracture		
									A0701	Capturing Problem		
									A072201	High Impedance		
									A090801	Signal Artifact/Noise		
							E2403	No Clinical Signs, Symptoms or Conditions	A040101	Fracture		
									A0701	Capturing Problem		
									A072201	High Impedance		
									A090801	Signal Artifact/Noise		
							F12	Serious Injury/ Illness/ Impairment	A040101	Fracture		
									A0701	Capturing Problem		
									A072201	High Impedance		
									A090801	Signal Artifact/Noise		
							F23	Unexpected Medical Intervention	A040101	Fracture		
									A0701	Capturing Problem		
									A072201	High Impedance		
									A090801	Signal Artifact/Noise		
6586	2022	All other reportable incidents	37265	Dual-chamber implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E060103	Atrial Flutter	A071202	Inaccurate Synchronization		
							E2403	No Clinical Signs, Symptoms or Conditions	A071202	Inaccurate Synchronization		
							F15	Recognized Device or Procedural Complication	A071202	Inaccurate Synchronization		
6587	2022	All other reportable incidents	17882	Physiologic-monitoring defibrillation system	CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A0719	Unexpected Shutdown		
							F26	No Health Consequences or Impact	A0719	Unexpected Shutdown		
6588	2022	All other reportable incidents	17882	Physiologic-monitoring defibrillation system	CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070903	Failure to Analyze Signal		
							F26	No Health Consequences or Impact	A070903	Failure to Analyze Signal		
6589	2021	All other reportable incidents	35852	Single-chamber implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E2104	Electric Shock	A070909	Oversensing		
									A071302	Inappropriate/Inadequate Shock/Stimulation		
							E2403	No Clinical Signs, Symptoms or Conditions	A070909	Oversensing		
									A071302	Inappropriate/Inadequate Shock/Stimulation		
							F08	Hospitalization or Prolonged Hospitalization	A070909	Oversensing		
									A071302	Inappropriate/Inadequate Shock/Stimulation		
							F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A070909	Oversensing		
									A071302	Inappropriate/Inadequate Shock/Stimulation		
6590	2021	All other reportable incidents	36241	Pacemaker/defibrillator lead	CT1684	Cardiac pacemakers and associated devices	E2104	Electric Shock	A070909	Oversensing		
									A071302	Inappropriate/Inadequate Shock/Stimulation		
									A090801	Signal Artifact/Noise		
							E2403	No Clinical Signs, Symptoms or Conditions	A070909	Oversensing		
									A071302	Inappropriate/Inadequate Shock/Stimulation		
									A090801	Signal Artifact/Noise		
							F08	Hospitalization or Prolonged Hospitalization	A070909	Oversensing		
									A071302	Inappropriate/Inadequate Shock/Stimulation		
									A090801	Signal Artifact/Noise		
							F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A070909	Oversensing		
									A071302	Inappropriate/Inadequate Shock/Stimulation		
									A090801	Signal Artifact/Noise		
6591	2022	All other reportable incidents	35852	Single-chamber implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem	28/04/2022	28/04/2022
									A070504	Premature Discharge of Battery	28/04/2022	28/04/2022
							F08	Hospitalization or Prolonged Hospitalization	A0705	Battery Problem	28/04/2022	28/04/2022
									A070504	Premature Discharge of Battery	28/04/2022	28/04/2022
							F1905	Device Revision or Replacement	A0705	Battery Problem	28/04/2022	28/04/2022
									A070504	Premature Discharge of Battery	28/04/2022	28/04/2022
6592	2022	All other reportable incidents	47263	Cardiac resynchronization therapy implantable pacemaker	CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A0908	Incorrect, Inadequate or Imprecise Result or Readings	07/04/2022	07/04/2022
							F08	Hospitalization or Prolonged Hospitalization	A0908	Incorrect, Inadequate or Imprecise Result or Readings	07/04/2022	07/04/2022
							F1905	Device Revision or Replacement	A0908	Incorrect, Inadequate or Imprecise Result or Readings	07/04/2022	07/04/2022
					CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A0908	Incorrect, Inadequate or Imprecise Result or Readings	07/04/2022	07/04/2022
							F08	Hospitalization or Prolonged Hospitalization	A0908	Incorrect, Inadequate or Imprecise Result or Readings	07/04/2022	07/04/2022
							F1905	Device Revision or Replacement	A0908	Incorrect, Inadequate or Imprecise Result or Readings	07/04/2022	07/04/2022
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A0908	Incorrect, Inadequate or Imprecise Result or Readings	07/04/2022	07/04/2022
							F08	Hospitalization or Prolonged Hospitalization	A0908	Incorrect, Inadequate or Imprecise Result or Readings	07/04/2022	07/04/2022
							F1905	Device Revision or Replacement	A0908	Incorrect, Inadequate or Imprecise Result or Readings	07/04/2022	07/04/2022
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A0908	Incorrect, Inadequate or Imprecise Result or Readings	07/04/2022	07/04/2022
							F08	Hospitalization or Prolonged Hospitalization	A0908	Incorrect, Inadequate or Imprecise Result or Readings	07/04/2022	07/04/2022
							F1905	Device Revision or Replacement	A0908	Incorrect, Inadequate or Imprecise Result or Readings	07/04/2022	07/04/2022
6593	2022	All other reportable incidents	35852	Single-chamber implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem	05/07/2022	05/07/2022
									A070504	Premature Discharge of Battery	05/07/2022	05/07/2022
							F08	Hospitalization or Prolonged Hospitalization	A0705	Battery Problem	05/07/2022	05/07/2022
									A070504	Premature Discharge of Battery	05/07/2022	05/07/2022
							F1905	Device Revision or Replacement	A0705	Battery Problem	05/07/2022	05/07/2022
									A070504	Premature Discharge of Battery	05/07/2022	05/07/2022
6594	2021	Death	35223	Endocardial/interventricular septal pacing lead	CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A0401	Break		
									A090801	Signal Artifact/Noise		

							F02	Death	A0401	Break		
									A090801	Signal Artifact/Noise		
6595	2021	All other reportable incidents	47265	Dual-chamber implantable pacemaker, rate-responsive	CT141	Electrodes/Leads and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070909	Oversensing		
									A070910	Undersensing		
							F26	No Health Consequences or Impact	A070909	Oversensing		
									A070910	Undersensing		
					CT145	Cardiovascular devices	E2403	No Clinical Signs, Symptoms or Conditions	A070909	Oversensing		
									A070910	Undersensing		
							F26	No Health Consequences or Impact	A070909	Oversensing		
									A070910	Undersensing		
					CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070909	Oversensing		
									A070910	Undersensing		
							F26	No Health Consequences or Impact	A070909	Oversensing		
									A070910	Undersensing		
					CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A070909	Oversensing		
									A070910	Undersensing		
							F26	No Health Consequences or Impact	A070909	Oversensing		
									A070910	Undersensing		
					CT233	Surgical	E2403	No Clinical Signs, Symptoms or Conditions	A070909	Oversensing		
									A070910	Undersensing		
							F26	No Health Consequences or Impact	A070909	Oversensing		
									A070910	Undersensing		
					CT2406	Implantable	E2403	No Clinical Signs, Symptoms or Conditions	A070909	Oversensing		
									A070910	Undersensing		
							F26	No Health Consequences or Impact	A070909	Oversensing		
									A070910	Undersensing		
					CT2407	Active implantable	E2403	No Clinical Signs, Symptoms or Conditions	A070909	Oversensing		
									A070910	Undersensing		
							F26	No Health Consequences or Impact	A070909	Oversensing		
									A070910	Undersensing		
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A070909	Oversensing		
									A070910	Undersensing		
							F26	No Health Consequences or Impact	A070909	Oversensing		
									A070910	Undersensing		
					CT2727	Active implant devices/ancillaries	E2403	No Clinical Signs, Symptoms or Conditions	A070909	Oversensing		
									A070910	Undersensing		
							F26	No Health Consequences or Impact	A070909	Oversensing		
									A070910	Undersensing		
					CT2750	Implantable cardiac electrical stimulators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070909	Oversensing		
									A070910	Undersensing		
							F26	No Health Consequences or Impact	A070909	Oversensing		
									A070910	Undersensing		
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A070909	Oversensing		
									A070910	Undersensing		
							F26	No Health Consequences or Impact	A070909	Oversensing		
									A070910	Undersensing		
					CT2985	Active	E2403	No Clinical Signs, Symptoms or Conditions	A070909	Oversensing		
									A070910	Undersensing		
							F26	No Health Consequences or Impact	A070909	Oversensing		
									A070910	Undersensing		
					CT3001	Implantable metal	E2403	No Clinical Signs, Symptoms or Conditions	A070909	Oversensing		
									A070910	Undersensing		
							F26	No Health Consequences or Impact	A070909	Oversensing		
									A070910	Undersensing		
					CT3002	Implantable synthetic polymer	E2403	No Clinical Signs, Symptoms or Conditions	A070909	Oversensing		
									A070910	Undersensing		
							F26	No Health Consequences or Impact	A070909	Oversensing		
									A070910	Undersensing		
					CT3006	Implantable inorganic material	E2403	No Clinical Signs, Symptoms or Conditions	A070909	Oversensing		
									A070910	Undersensing		
							F26	No Health Consequences or Impact	A070909	Oversensing		
									A070910	Undersensing		
					CT314	Electrically powered	E2403	No Clinical Signs, Symptoms or Conditions	A070909	Oversensing		
									A070910	Undersensing		
							F26	No Health Consequences or Impact	A070909	Oversensing		
									A070910	Undersensing		
					CT321	Long-term surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A070909	Oversensing		
									A070910	Undersensing		
							F26	No Health Consequences or Impact	A070909	Oversensing		
									A070910	Undersensing		
					CT449	Generators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070909	Oversensing		
									A070910	Undersensing		
							F26	No Health Consequences or Impact	A070909	Oversensing		
									A070910	Undersensing		
					CT981	Single use	E2403	No Clinical Signs, Symptoms or Conditions	A070909	Oversensing		
									A070910	Undersensing		
							F26	No Health Consequences or Impact	A070909	Oversensing		
									A070910	Undersensing		
					CT983	Surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A070909	Oversensing		
									A070910	Undersensing		
							F26	No Health Consequences or Impact	A070909	Oversensing		
									A070910	Undersensing		
6596	2021	All other reportable incidents	47265	Dual-chamber implantable pacemaker, rate-responsive	CT141	Electrodes/Leads and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	31/08/2021	31/08/2021
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	31/08/2021	31/08/2021
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	31/08/2021	31/08/2021
					CT145	Cardiovascular devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	31/08/2021	31/08/2021
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	31/08/2021	31/08/2021
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	31/08/2021	31/08/2021

					CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	31/08/2021	31/08/2021
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	31/08/2021	31/08/2021
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	31/08/2021	31/08/2021
					CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	31/08/2021	31/08/2021
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	31/08/2021	31/08/2021
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	31/08/2021	31/08/2021
					CT233	Surgical	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	31/08/2021	31/08/2021
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	31/08/2021	31/08/2021
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	31/08/2021	31/08/2021
					CT2406	Implantable	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	31/08/2021	31/08/2021
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	31/08/2021	31/08/2021
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	31/08/2021	31/08/2021
					CT2407	Active implantable	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	31/08/2021	31/08/2021
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	31/08/2021	31/08/2021
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	31/08/2021	31/08/2021
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	31/08/2021	31/08/2021
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	31/08/2021	31/08/2021
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	31/08/2021	31/08/2021
					CT2727	Active implant devices/ancillaries	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	31/08/2021	31/08/2021
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	31/08/2021	31/08/2021
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	31/08/2021	31/08/2021
					CT2750	Implantable cardiac electrical stimulators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	31/08/2021	31/08/2021
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	31/08/2021	31/08/2021
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	31/08/2021	31/08/2021
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	31/08/2021	31/08/2021
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	31/08/2021	31/08/2021
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	31/08/2021	31/08/2021
					CT2985	Active	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	31/08/2021	31/08/2021
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	31/08/2021	31/08/2021
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	31/08/2021	31/08/2021
					CT3001	Implantable metal	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	31/08/2021	31/08/2021
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	31/08/2021	31/08/2021
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	31/08/2021	31/08/2021
					CT3002	Implantable synthetic polymer	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	31/08/2021	31/08/2021
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	31/08/2021	31/08/2021
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	31/08/2021	31/08/2021
					CT3006	Implantable inorganic material	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	31/08/2021	31/08/2021
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	31/08/2021	31/08/2021
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	31/08/2021	31/08/2021
					CT314	Electrically powered	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	31/08/2021	31/08/2021
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	31/08/2021	31/08/2021
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	31/08/2021	31/08/2021
					CT321	Long-term surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	31/08/2021	31/08/2021
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	31/08/2021	31/08/2021
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	31/08/2021	31/08/2021
					CT449	Generators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	31/08/2021	31/08/2021
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	31/08/2021	31/08/2021
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	31/08/2021	31/08/2021
					CT981	Single use	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	31/08/2021	31/08/2021
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	31/08/2021	31/08/2021
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	31/08/2021	31/08/2021
					CT983	Surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	31/08/2021	31/08/2021
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	31/08/2021	31/08/2021
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	31/08/2021	31/08/2021
6597	2021	All other reportable incidents	35852	Single-chamber implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A070909	Oversensing		
							F26	No Health Consequences or Impact	A070909	Oversensing		
6598	2021	All other reportable incidents	35852	Single-chamber implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem	18/10/2021	18/10/2021
									A070504	Premature Discharge of Battery	18/10/2021	18/10/2021
							F08	Hospitalization or Prolonged Hospitalization	A0705	Battery Problem	18/10/2021	18/10/2021
									A070504	Premature Discharge of Battery	18/10/2021	18/10/2021
							F1905	Device Revision or Replacement	A0705	Battery Problem	18/10/2021	18/10/2021
									A070504	Premature Discharge of Battery	18/10/2021	18/10/2021
6599	2022	All other reportable incidents	47267	Single-chamber implantable pacemaker, rate-responsive	CT141	Electrodes/Leads and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	13/06/2022	13/06/2022
									A071102	Failure to Interrogate	13/06/2022	13/06/2022
							F1901	Additional Surgery	A070504	Premature Discharge of Battery	13/06/2022	13/06/2022
									A071102	Failure to Interrogate	13/06/2022	13/06/2022
					CT145	Cardiovascular devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	13/06/2022	13/06/2022
									A071102	Failure to Interrogate	13/06/2022	13/06/2022
							F1901	Additional Surgery	A070504	Premature Discharge of Battery	13/06/2022	13/06/2022
									A071102	Failure to Interrogate	13/06/2022	13/06/2022
					CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	13/06/2022	13/06/2022
									A071102	Failure to Interrogate	13/06/2022	13/06/2022
							F1901	Additional Surgery	A070504	Premature Discharge of Battery	13/06/2022	13/06/2022
									A071102	Failure to Interrogate	13/06/2022	13/06/2022
					CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	13/06/2022	13/06/2022
									A071102	Failure to Interrogate	13/06/2022	13/06/2022
							F1901	Additional Surgery	A070504	Premature Discharge of Battery	13/06/2022	13/06/2022
									A071102	Failure to Interrogate	13/06/2022	13/06/2022
					CT233	Surgical	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	13/06/2022	13/06/2022
									A071102	Failure to Interrogate	13/06/2022	13/06/2022
							F1901	Additional Surgery	A070504	Premature Discharge of Battery	13/06/2022	13/06/2022
									A071102	Failure to Interrogate	13/06/2022	13/06/2022
					CT2406	Implantable	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	13/06/2022	13/06/2022
									A071102	Failure to Interrogate	13/06/2022	13/06/2022
							F1901	Additional Surgery	A070504	Premature Discharge of Battery	13/06/2022	13/06/2022
									A071102	Failure to Interrogate	13/06/2022	13/06/2022
					CT2407	Active implantable	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	13/06/2022	13/06/2022
									A071102	Failure to Interrogate	13/06/2022	13/06/2022

							F1901	Additional Surgery	A070504	Premature Discharge of Battery	13/06/2022	13/06/2022
									A071102	Failure to Interrogate	13/06/2022	13/06/2022
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	13/06/2022	13/06/2022
									A071102	Failure to Interrogate	13/06/2022	13/06/2022
							F1901	Additional Surgery	A070504	Premature Discharge of Battery	13/06/2022	13/06/2022
									A071102	Failure to Interrogate	13/06/2022	13/06/2022
					CT2727	Active implant devices/ancillaries	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	13/06/2022	13/06/2022
									A071102	Failure to Interrogate	13/06/2022	13/06/2022
							F1901	Additional Surgery	A070504	Premature Discharge of Battery	13/06/2022	13/06/2022
									A071102	Failure to Interrogate	13/06/2022	13/06/2022
					CT2750	Implantable cardiac electrical stimulators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	13/06/2022	13/06/2022
									A071102	Failure to Interrogate	13/06/2022	13/06/2022
							F1901	Additional Surgery	A070504	Premature Discharge of Battery	13/06/2022	13/06/2022
									A071102	Failure to Interrogate	13/06/2022	13/06/2022
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	13/06/2022	13/06/2022
									A071102	Failure to Interrogate	13/06/2022	13/06/2022
							F1901	Additional Surgery	A070504	Premature Discharge of Battery	13/06/2022	13/06/2022
									A071102	Failure to Interrogate	13/06/2022	13/06/2022
					CT2985	Active	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	13/06/2022	13/06/2022
									A071102	Failure to Interrogate	13/06/2022	13/06/2022
							F1901	Additional Surgery	A070504	Premature Discharge of Battery	13/06/2022	13/06/2022
									A071102	Failure to Interrogate	13/06/2022	13/06/2022
					CT3001	Implantable metal	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	13/06/2022	13/06/2022
									A071102	Failure to Interrogate	13/06/2022	13/06/2022
							F1901	Additional Surgery	A070504	Premature Discharge of Battery	13/06/2022	13/06/2022
									A071102	Failure to Interrogate	13/06/2022	13/06/2022
					CT3002	Implantable synthetic polymer	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	13/06/2022	13/06/2022
									A071102	Failure to Interrogate	13/06/2022	13/06/2022
							F1901	Additional Surgery	A070504	Premature Discharge of Battery	13/06/2022	13/06/2022
									A071102	Failure to Interrogate	13/06/2022	13/06/2022
					CT3006	Implantable inorganic material	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	13/06/2022	13/06/2022
									A071102	Failure to Interrogate	13/06/2022	13/06/2022
							F1901	Additional Surgery	A070504	Premature Discharge of Battery	13/06/2022	13/06/2022
									A071102	Failure to Interrogate	13/06/2022	13/06/2022
					CT314	Electrically powered	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	13/06/2022	13/06/2022
									A071102	Failure to Interrogate	13/06/2022	13/06/2022
							F1901	Additional Surgery	A070504	Premature Discharge of Battery	13/06/2022	13/06/2022
									A071102	Failure to Interrogate	13/06/2022	13/06/2022
					CT321	Long-term surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	13/06/2022	13/06/2022
									A071102	Failure to Interrogate	13/06/2022	13/06/2022
							F1901	Additional Surgery	A070504	Premature Discharge of Battery	13/06/2022	13/06/2022
									A071102	Failure to Interrogate	13/06/2022	13/06/2022
					CT449	Generators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	13/06/2022	13/06/2022
									A071102	Failure to Interrogate	13/06/2022	13/06/2022
							F1901	Additional Surgery	A070504	Premature Discharge of Battery	13/06/2022	13/06/2022
									A071102	Failure to Interrogate	13/06/2022	13/06/2022
					CT981	Single use	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	13/06/2022	13/06/2022
									A071102	Failure to Interrogate	13/06/2022	13/06/2022
							F1901	Additional Surgery	A070504	Premature Discharge of Battery	13/06/2022	13/06/2022
									A071102	Failure to Interrogate	13/06/2022	13/06/2022
					CT983	Surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	13/06/2022	13/06/2022
									A071102	Failure to Interrogate	13/06/2022	13/06/2022
							F1901	Additional Surgery	A070504	Premature Discharge of Battery	13/06/2022	13/06/2022
									A071102	Failure to Interrogate	13/06/2022	13/06/2022
6600	2021	All other reportable incidents	35223	Endocardial/interventricular septal pacing lead	CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A0401	Break	04/11/2021	04/11/2021
									A040101	Fracture	04/11/2021	04/11/2021
									A090801	Signal Artifact/Noise	04/11/2021	04/11/2021
							F26	No Health Consequences or Impact	A0401	Break	04/11/2021	04/11/2021
									A040101	Fracture	04/11/2021	04/11/2021
									A090801	Signal Artifact/Noise	04/11/2021	04/11/2021
6601	2022	All other reportable incidents	17882	Physiologic-monitoring defibrillation system	CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A0908	Incorrect, Inadequate or Imprecise Result or Readings		
							F27	Problem Identified During Non-Clinical Procedure	A0908	Incorrect, Inadequate or Imprecise Result or Readings		
6602	2022	All other reportable incidents	17882	Physiologic-monitoring defibrillation system	CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A1603	Failure of Device to Self-Test		
							F27	Problem Identified During Non-Clinical Procedure	A1603	Failure of Device to Self-Test		
6603	2022	All other reportable incidents	17882	Physiologic-monitoring defibrillation system	CT1684	Cardiac pacemakers and associated devices	E2401	Insufficient Information	A090207	No Visual Prompts/Feedback		
							E2403	No Clinical Signs, Symptoms or Conditions	A090207	No Visual Prompts/Feedback		
							F26	No Health Consequences or Impact	A090207	No Visual Prompts/Feedback		
6604	2022	All other reportable incidents	17882	Physiologic-monitoring defibrillation system	CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070908	Failure to Sense		
							F26	No Health Consequences or Impact	A070908	Failure to Sense		
6605	2021	All other reportable incidents	47263	Cardiac resynchronization therapy implantable pacemaker	CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A0908	Incorrect, Inadequate or Imprecise Result or Readings	17/09/2021	17/09/2021
							F08	Hospitalization or Prolonged Hospitalization	A0908	Incorrect, Inadequate or Imprecise Result or Readings	17/09/2021	17/09/2021
							F1905	Device Revision or Replacement	A0908	Incorrect, Inadequate or Imprecise Result or Readings	17/09/2021	17/09/2021
					CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A0908	Incorrect, Inadequate or Imprecise Result or Readings	17/09/2021	17/09/2021
							F08	Hospitalization or Prolonged Hospitalization	A0908	Incorrect, Inadequate or Imprecise Result or Readings	17/09/2021	17/09/2021
							F1905	Device Revision or Replacement	A0908	Incorrect, Inadequate or Imprecise Result or Readings	17/09/2021	17/09/2021
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A0908	Incorrect, Inadequate or Imprecise Result or Readings	17/09/2021	17/09/2021
							F08	Hospitalization or Prolonged Hospitalization	A0908	Incorrect, Inadequate or Imprecise Result or Readings	17/09/2021	17/09/2021
							F1905	Device Revision or Replacement	A0908	Incorrect, Inadequate or Imprecise Result or Readings	17/09/2021	17/09/2021
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A0908	Incorrect, Inadequate or Imprecise Result or Readings	17/09/2021	17/09/2021

							F08	Hospitalization or Prolonged Hospitalization	A0908	Incorrect, Inadequate or Imprecise Result or Readings	17/09/2021	17/09/2021
							F1905	Device Revision or Replacement	A0908	Incorrect, Inadequate or Imprecise Result or Readings	17/09/2021	17/09/2021
6606	2021	All other reportable incidents	47265	Dual-chamber implantable pacemaker, rate-responsive	CT141	Electrodes/Leads and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	10/09/2021	10/09/2021
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	10/09/2021	10/09/2021
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	10/09/2021	10/09/2021
					CT145	Cardiovascular devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	10/09/2021	10/09/2021
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	10/09/2021	10/09/2021
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	10/09/2021	10/09/2021
					CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	10/09/2021	10/09/2021
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	10/09/2021	10/09/2021
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	10/09/2021	10/09/2021
					CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	10/09/2021	10/09/2021
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	10/09/2021	10/09/2021
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	10/09/2021	10/09/2021
					CT233	Surgical	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	10/09/2021	10/09/2021
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	10/09/2021	10/09/2021
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	10/09/2021	10/09/2021
					CT2406	Implantable	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	10/09/2021	10/09/2021
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	10/09/2021	10/09/2021
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	10/09/2021	10/09/2021
					CT2407	Active implantable	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	10/09/2021	10/09/2021
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	10/09/2021	10/09/2021
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	10/09/2021	10/09/2021
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	10/09/2021	10/09/2021
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	10/09/2021	10/09/2021
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	10/09/2021	10/09/2021
					CT2727	Active implant devices/ancillaries	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	10/09/2021	10/09/2021
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	10/09/2021	10/09/2021
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	10/09/2021	10/09/2021
					CT2750	Implantable cardiac electrical stimulators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	10/09/2021	10/09/2021
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	10/09/2021	10/09/2021
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	10/09/2021	10/09/2021
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	10/09/2021	10/09/2021
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	10/09/2021	10/09/2021
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	10/09/2021	10/09/2021
					CT2985	Active	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	10/09/2021	10/09/2021
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	10/09/2021	10/09/2021
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	10/09/2021	10/09/2021
					CT3001	Implantable metal	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	10/09/2021	10/09/2021
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	10/09/2021	10/09/2021
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	10/09/2021	10/09/2021
					CT3002	Implantable synthetic polymer	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	10/09/2021	10/09/2021
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	10/09/2021	10/09/2021
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	10/09/2021	10/09/2021
					CT3006	Implantable inorganic material	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	10/09/2021	10/09/2021
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	10/09/2021	10/09/2021
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	10/09/2021	10/09/2021
					CT314	Electrically powered	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	10/09/2021	10/09/2021
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	10/09/2021	10/09/2021
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	10/09/2021	10/09/2021
					CT321	Long-term surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	10/09/2021	10/09/2021
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	10/09/2021	10/09/2021
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	10/09/2021	10/09/2021
					CT449	Generators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	10/09/2021	10/09/2021
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	10/09/2021	10/09/2021
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	10/09/2021	10/09/2021
					CT981	Single use	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	10/09/2021	10/09/2021
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	10/09/2021	10/09/2021
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	10/09/2021	10/09/2021
					CT983	Surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	10/09/2021	10/09/2021
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	10/09/2021	10/09/2021
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	10/09/2021	10/09/2021
6607	2022	All other reportable incidents	47265	Dual-chamber implantable pacemaker, rate-responsive	CT141	Electrodes/Leads and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	01/05/2022	01/05/2022
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	01/05/2022	01/05/2022
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	01/05/2022	01/05/2022
					CT145	Cardiovascular devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	01/05/2022	01/05/2022
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	01/05/2022	01/05/2022
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	01/05/2022	01/05/2022
					CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	01/05/2022	01/05/2022
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	01/05/2022	01/05/2022
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	01/05/2022	01/05/2022
					CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	01/05/2022	01/05/2022
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	01/05/2022	01/05/2022
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	01/05/2022	01/05/2022
					CT233	Surgical	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	01/05/2022	01/05/2022
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	01/05/2022	01/05/2022
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	01/05/2022	01/05/2022
					CT2406	Implantable	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	01/05/2022	01/05/2022
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	01/05/2022	01/05/2022
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	01/05/2022	01/05/2022
					CT2407	Active implantable	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	01/05/2022	01/05/2022
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	01/05/2022	01/05/2022
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	01/05/2022	01/05/2022
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	01/05/2022	01/05/2022
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	01/05/2022	01/05/2022
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	01/05/2022	01/05/2022

					CT2727	Active implant devices/ancillaries	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	01/05/2022	01/05/2022
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	01/05/2022	01/05/2022
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	01/05/2022	01/05/2022
					CT2750	Implantable cardiac electrical stimulators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	01/05/2022	01/05/2022
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	01/05/2022	01/05/2022
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	01/05/2022	01/05/2022
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	01/05/2022	01/05/2022
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	01/05/2022	01/05/2022
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	01/05/2022	01/05/2022
					CT2985	Active	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	01/05/2022	01/05/2022
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	01/05/2022	01/05/2022
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	01/05/2022	01/05/2022
					CT3001	Implantable metal	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	01/05/2022	01/05/2022
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	01/05/2022	01/05/2022
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	01/05/2022	01/05/2022
					CT3002	Implantable synthetic polymer	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	01/05/2022	01/05/2022
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	01/05/2022	01/05/2022
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	01/05/2022	01/05/2022
					CT3006	Implantable inorganic material	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	01/05/2022	01/05/2022
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	01/05/2022	01/05/2022
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	01/05/2022	01/05/2022
					CT314	Electrically powered	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	01/05/2022	01/05/2022
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	01/05/2022	01/05/2022
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	01/05/2022	01/05/2022
					CT321	Long-term surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	01/05/2022	01/05/2022
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	01/05/2022	01/05/2022
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	01/05/2022	01/05/2022
					CT449	Generators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	01/05/2022	01/05/2022
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	01/05/2022	01/05/2022
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	01/05/2022	01/05/2022
					CT981	Single use	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	01/05/2022	01/05/2022
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	01/05/2022	01/05/2022
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	01/05/2022	01/05/2022
					CT983	Surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	01/05/2022	01/05/2022
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	01/05/2022	01/05/2022
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	01/05/2022	01/05/2022
6608	2021	All other reportable incidents	36241	Pacemaker/defibrillator lead	CT1684	Cardiac pacemakers and associated devices	E0112	Dizziness	A070101	Failure to Capture		
									A070102	High Capture Threshold		
							E2403	No Clinical Signs, Symptoms or Conditions	A070101	Failure to Capture		
									A070102	High Capture Threshold		
							F08	Hospitalization or Prolonged Hospitalization	A070101	Failure to Capture		
									A070102	High Capture Threshold		
							F1904	Device Repositioning	A070101	Failure to Capture		
									A070102	High Capture Threshold		
							F2203	Imaging Required	A070101	Failure to Capture		
									A070102	High Capture Threshold		
6609	2021	All other reportable incidents	36241	Pacemaker/defibrillator lead	CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A072201	High Impedance		
							F26	No Health Consequences or Impact	A072201	High Impedance		
6610	2021	All other reportable incidents	47265	Dual-chamber implantable pacemaker, rate-responsive	CT141	Electrodes/Leads and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem	15/09/2021	15/09/2021
									A070504	Premature Discharge of Battery	15/09/2021	15/09/2021
							F08	Hospitalization or Prolonged Hospitalization	A0705	Battery Problem	15/09/2021	15/09/2021
									A070504	Premature Discharge of Battery	15/09/2021	15/09/2021
							F1905	Device Revision or Replacement	A0705	Battery Problem	15/09/2021	15/09/2021
									A070504	Premature Discharge of Battery	15/09/2021	15/09/2021
					CT145	Cardiovascular devices	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem	15/09/2021	15/09/2021
									A070504	Premature Discharge of Battery	15/09/2021	15/09/2021
							F08	Hospitalization or Prolonged Hospitalization	A0705	Battery Problem	15/09/2021	15/09/2021
									A070504	Premature Discharge of Battery	15/09/2021	15/09/2021
							F1905	Device Revision or Replacement	A0705	Battery Problem	15/09/2021	15/09/2021
									A070504	Premature Discharge of Battery	15/09/2021	15/09/2021
					CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem	15/09/2021	15/09/2021
									A070504	Premature Discharge of Battery	15/09/2021	15/09/2021
							F08	Hospitalization or Prolonged Hospitalization	A0705	Battery Problem	15/09/2021	15/09/2021
									A070504	Premature Discharge of Battery	15/09/2021	15/09/2021
							F1905	Device Revision or Replacement	A0705	Battery Problem	15/09/2021	15/09/2021
									A070504	Premature Discharge of Battery	15/09/2021	15/09/2021
					CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem	15/09/2021	15/09/2021
									A070504	Premature Discharge of Battery	15/09/2021	15/09/2021
							F08	Hospitalization or Prolonged Hospitalization	A0705	Battery Problem	15/09/2021	15/09/2021
									A070504	Premature Discharge of Battery	15/09/2021	15/09/2021
							F1905	Device Revision or Replacement	A0705	Battery Problem	15/09/2021	15/09/2021
									A070504	Premature Discharge of Battery	15/09/2021	15/09/2021
					CT233	Surgical	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem	15/09/2021	15/09/2021
									A070504	Premature Discharge of Battery	15/09/2021	15/09/2021
							F08	Hospitalization or Prolonged Hospitalization	A0705	Battery Problem	15/09/2021	15/09/2021
									A070504	Premature Discharge of Battery	15/09/2021	15/09/2021
							F1905	Device Revision or Replacement	A0705	Battery Problem	15/09/2021	15/09/2021
									A070504	Premature Discharge of Battery	15/09/2021	15/09/2021
					CT2406	Implantable	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem	15/09/2021	15/09/2021
									A070504	Premature Discharge of Battery	15/09/2021	15/09/2021
							F08	Hospitalization or Prolonged Hospitalization	A0705	Battery Problem	15/09/2021	15/09/2021
									A070504	Premature Discharge of Battery	15/09/2021	15/09/2021
							F1905	Device Revision or Replacement	A0705	Battery Problem	15/09/2021	15/09/2021
									A070504	Premature Discharge of Battery	15/09/2021	15/09/2021
					CT2407	Active implantable	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem	15/09/2021	15/09/2021
									A070504	Premature Discharge of Battery	15/09/2021	15/09/2021
							F08	Hospitalization or Prolonged Hospitalization	A0705	Battery Problem	15/09/2021	15/09/2021
									A070504	Premature Discharge of Battery	15/09/2021	15/09/2021

							F1905	Device Revision or Replacement	A0705	Battery Problem	15/09/2021	15/09/2021
									A070504	Premature Discharge of Battery	15/09/2021	15/09/2021
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem	15/09/2021	15/09/2021
									A070504	Premature Discharge of Battery	15/09/2021	15/09/2021
							F08	Hospitalization or Prolonged Hospitalization	A0705	Battery Problem	15/09/2021	15/09/2021
									A070504	Premature Discharge of Battery	15/09/2021	15/09/2021
							F1905	Device Revision or Replacement	A0705	Battery Problem	15/09/2021	15/09/2021
									A070504	Premature Discharge of Battery	15/09/2021	15/09/2021
					CT2727	Active implant devices/ancillaries	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem	15/09/2021	15/09/2021
									A070504	Premature Discharge of Battery	15/09/2021	15/09/2021
							F08	Hospitalization or Prolonged Hospitalization	A0705	Battery Problem	15/09/2021	15/09/2021
									A070504	Premature Discharge of Battery	15/09/2021	15/09/2021
							F1905	Device Revision or Replacement	A0705	Battery Problem	15/09/2021	15/09/2021
									A070504	Premature Discharge of Battery	15/09/2021	15/09/2021
					CT2750	Implantable cardiac electrical stimulators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem	15/09/2021	15/09/2021
									A070504	Premature Discharge of Battery	15/09/2021	15/09/2021
							F08	Hospitalization or Prolonged Hospitalization	A0705	Battery Problem	15/09/2021	15/09/2021
									A070504	Premature Discharge of Battery	15/09/2021	15/09/2021
							F1905	Device Revision or Replacement	A0705	Battery Problem	15/09/2021	15/09/2021
									A070504	Premature Discharge of Battery	15/09/2021	15/09/2021
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem	15/09/2021	15/09/2021
									A070504	Premature Discharge of Battery	15/09/2021	15/09/2021
							F08	Hospitalization or Prolonged Hospitalization	A0705	Battery Problem	15/09/2021	15/09/2021
									A070504	Premature Discharge of Battery	15/09/2021	15/09/2021
							F1905	Device Revision or Replacement	A0705	Battery Problem	15/09/2021	15/09/2021
									A070504	Premature Discharge of Battery	15/09/2021	15/09/2021
					CT2985	Active	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem	15/09/2021	15/09/2021
									A070504	Premature Discharge of Battery	15/09/2021	15/09/2021
							F08	Hospitalization or Prolonged Hospitalization	A0705	Battery Problem	15/09/2021	15/09/2021
									A070504	Premature Discharge of Battery	15/09/2021	15/09/2021
							F1905	Device Revision or Replacement	A0705	Battery Problem	15/09/2021	15/09/2021
									A070504	Premature Discharge of Battery	15/09/2021	15/09/2021
					CT3001	Implantable metal	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem	15/09/2021	15/09/2021
									A070504	Premature Discharge of Battery	15/09/2021	15/09/2021
							F08	Hospitalization or Prolonged Hospitalization	A0705	Battery Problem	15/09/2021	15/09/2021
									A070504	Premature Discharge of Battery	15/09/2021	15/09/2021
							F1905	Device Revision or Replacement	A0705	Battery Problem	15/09/2021	15/09/2021
									A070504	Premature Discharge of Battery	15/09/2021	15/09/2021
					CT3002	Implantable synthetic polymer	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem	15/09/2021	15/09/2021
									A070504	Premature Discharge of Battery	15/09/2021	15/09/2021
							F08	Hospitalization or Prolonged Hospitalization	A0705	Battery Problem	15/09/2021	15/09/2021
									A070504	Premature Discharge of Battery	15/09/2021	15/09/2021
							F1905	Device Revision or Replacement	A0705	Battery Problem	15/09/2021	15/09/2021
									A070504	Premature Discharge of Battery	15/09/2021	15/09/2021
					CT3006	Implantable inorganic material	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem	15/09/2021	15/09/2021
									A070504	Premature Discharge of Battery	15/09/2021	15/09/2021
							F08	Hospitalization or Prolonged Hospitalization	A0705	Battery Problem	15/09/2021	15/09/2021
									A070504	Premature Discharge of Battery	15/09/2021	15/09/2021
							F1905	Device Revision or Replacement	A0705	Battery Problem	15/09/2021	15/09/2021
									A070504	Premature Discharge of Battery	15/09/2021	15/09/2021
					CT314	Electrically powered	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem	15/09/2021	15/09/2021
									A070504	Premature Discharge of Battery	15/09/2021	15/09/2021
							F08	Hospitalization or Prolonged Hospitalization	A0705	Battery Problem	15/09/2021	15/09/2021
									A070504	Premature Discharge of Battery	15/09/2021	15/09/2021
							F1905	Device Revision or Replacement	A0705	Battery Problem	15/09/2021	15/09/2021
									A070504	Premature Discharge of Battery	15/09/2021	15/09/2021
					CT321	Long-term surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem	15/09/2021	15/09/2021
									A070504	Premature Discharge of Battery	15/09/2021	15/09/2021
							F08	Hospitalization or Prolonged Hospitalization	A0705	Battery Problem	15/09/2021	15/09/2021
									A070504	Premature Discharge of Battery	15/09/2021	15/09/2021
							F1905	Device Revision or Replacement	A0705	Battery Problem	15/09/2021	15/09/2021
									A070504	Premature Discharge of Battery	15/09/2021	15/09/2021
					CT449	Generators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem	15/09/2021	15/09/2021
									A070504	Premature Discharge of Battery	15/09/2021	15/09/2021
							F08	Hospitalization or Prolonged Hospitalization	A0705	Battery Problem	15/09/2021	15/09/2021
									A070504	Premature Discharge of Battery	15/09/2021	15/09/2021
							F1905	Device Revision or Replacement	A0705	Battery Problem	15/09/2021	15/09/2021
									A070504	Premature Discharge of Battery	15/09/2021	15/09/2021
					CT981	Single use	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem	15/09/2021	15/09/2021
									A070504	Premature Discharge of Battery	15/09/2021	15/09/2021
							F08	Hospitalization or Prolonged Hospitalization	A0705	Battery Problem	15/09/2021	15/09/2021
									A070504	Premature Discharge of Battery	15/09/2021	15/09/2021
							F1905	Device Revision or Replacement	A0705	Battery Problem	15/09/2021	15/09/2021
									A070504	Premature Discharge of Battery	15/09/2021	15/09/2021
					CT983	Surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem	15/09/2021	15/09/2021
									A070504	Premature Discharge of Battery	15/09/2021	15/09/2021
							F08	Hospitalization or Prolonged Hospitalization	A0705	Battery Problem	15/09/2021	15/09/2021
									A070504	Premature Discharge of Battery	15/09/2021	15/09/2021
							F1905	Device Revision or Replacement	A0705	Battery Problem	15/09/2021	15/09/2021
									A070504	Premature Discharge of Battery	15/09/2021	15/09/2021
6611	2022	All other reportable incidents	37265	Dual-chamber implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	19/04/2022	19/04/2022
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	19/04/2022	19/04/2022
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	19/04/2022	19/04/2022
6612	2022	All other reportable incidents	37265	Dual-chamber implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	12/05/2022	12/05/2022
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	12/05/2022	12/05/2022
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	12/05/2022	12/05/2022
6613	2022	All other reportable incidents	47263	Cardiac resynchronization therapy implantable pacemaker	CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		

					CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
6614	2021	All other reportable incidents	47265	Dual-chamber implantable pacemaker, rate-responsive	CT141	Electrodes/Leads and associated devices	E060104	Bradycardia	A070909	Oversensing		
									A0712	Pacing Problem		
									A072202	Low Impedance		
									A090801	Signal Artifact/Noise		
									A1301	Failure to Read Input Signal		
							E2403	No Clinical Signs, Symptoms or Conditions	A070909	Oversensing		
									A0712	Pacing Problem		
									A072202	Low Impedance		
									A090801	Signal Artifact/Noise		
									A1301	Failure to Read Input Signal		
							F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A070909	Oversensing		
									A0712	Pacing Problem		
									A072202	Low Impedance		
									A090801	Signal Artifact/Noise		
									A1301	Failure to Read Input Signal		
							F15	Recognized Device or Procedural Complication	A070909	Oversensing		
									A0712	Pacing Problem		
									A072202	Low Impedance		
									A090801	Signal Artifact/Noise		
									A1301	Failure to Read Input Signal		
							F2203	Imaging Required	A070909	Oversensing		
									A0712	Pacing Problem		
									A072202	Low Impedance		
									A090801	Signal Artifact/Noise		
									A1301	Failure to Read Input Signal		
					CT145	Cardiovascular devices	E060104	Bradycardia	A070909	Oversensing		
									A0712	Pacing Problem		
									A072202	Low Impedance		
									A090801	Signal Artifact/Noise		
									A1301	Failure to Read Input Signal		
							E2403	No Clinical Signs, Symptoms or Conditions	A070909	Oversensing		
									A0712	Pacing Problem		
									A072202	Low Impedance		
									A090801	Signal Artifact/Noise		
									A1301	Failure to Read Input Signal		
							F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A070909	Oversensing		
									A0712	Pacing Problem		
									A072202	Low Impedance		
									A090801	Signal Artifact/Noise		
									A1301	Failure to Read Input Signal		
							F15	Recognized Device or Procedural Complication	A070909	Oversensing		
									A0712	Pacing Problem		
									A072202	Low Impedance		
									A090801	Signal Artifact/Noise		
									A1301	Failure to Read Input Signal		
							F2203	Imaging Required	A070909	Oversensing		
									A0712	Pacing Problem		
									A072202	Low Impedance		
									A090801	Signal Artifact/Noise		
									A1301	Failure to Read Input Signal		
					CT1684	Cardiac pacemakers and associated devices	E060104	Bradycardia	A070909	Oversensing		
									A0712	Pacing Problem		
									A072202	Low Impedance		
									A090801	Signal Artifact/Noise		
									A1301	Failure to Read Input Signal		
							E2403	No Clinical Signs, Symptoms or Conditions	A070909	Oversensing		
									A0712	Pacing Problem		
									A072202	Low Impedance		
									A090801	Signal Artifact/Noise		
									A1301	Failure to Read Input Signal		
							F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A070909	Oversensing		
									A0712	Pacing Problem		
									A072202	Low Impedance		
									A090801	Signal Artifact/Noise		
									A1301	Failure to Read Input Signal		
							F15	Recognized Device or Procedural Complication	A070909	Oversensing		
									A0712	Pacing Problem		
									A072202	Low Impedance		
									A090801	Signal Artifact/Noise		
									A1301	Failure to Read Input Signal		
							F2203	Imaging Required	A070909	Oversensing		
									A0712	Pacing Problem		
									A072202	Low Impedance		
									A090801	Signal Artifact/Noise		
									A1301	Failure to Read Input Signal		
					CT1685	Cardiac pacemakers	E060104	Bradycardia	A070909	Oversensing		
									A0712	Pacing Problem		
									A072202	Low Impedance		
									A090801	Signal Artifact/Noise		
									A1301	Failure to Read Input Signal		
							E2403	No Clinical Signs, Symptoms or Conditions	A070909	Oversensing		
									A0712	Pacing Problem		

									A072202	Low Impedance		
									A090801	Signal Artifact/Noise		
									A1301	Failure to Read Input Signal		
								F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A070909	Oversensing		
										A0712	Pacing Problem		
										A072202	Low Impedance		
										A090801	Signal Artifact/Noise		
										A1301	Failure to Read Input Signal		
								F15	Recognized Device or Procedural Complication	A070909	Oversensing		
										A0712	Pacing Problem		
										A072202	Low Impedance		
										A090801	Signal Artifact/Noise		
										A1301	Failure to Read Input Signal		
								F2203	Imaging Required	A070909	Oversensing		
										A0712	Pacing Problem		
										A072202	Low Impedance		
										A090801	Signal Artifact/Noise		
										A1301	Failure to Read Input Signal		
							CT233	Surgical	E060104	Bradycardia	A070909	Oversensing		
											A0712	Pacing Problem		
											A072202	Low Impedance		
											A090801	Signal Artifact/Noise		
											A1301	Failure to Read Input Signal		
									E2403	No Clinical Signs, Symptoms or Conditions	A070909	Oversensing		
											A0712	Pacing Problem		
											A072202	Low Impedance		
											A090801	Signal Artifact/Noise		
											A1301	Failure to Read Input Signal		
									F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A070909	Oversensing		
											A0712	Pacing Problem		
											A072202	Low Impedance		
											A090801	Signal Artifact/Noise		
											A1301	Failure to Read Input Signal		
									F15	Recognized Device or Procedural Complication	A070909	Oversensing		
											A0712	Pacing Problem		
											A072202	Low Impedance		
											A090801	Signal Artifact/Noise		
											A1301	Failure to Read Input Signal		
									F2203	Imaging Required	A070909	Oversensing		
											A0712	Pacing Problem		
											A072202	Low Impedance		
											A090801	Signal Artifact/Noise		
											A1301	Failure to Read Input Signal		
								CT2406	Implantable	E060104	Bradycardia	A070909	Oversensing		
												A0712	Pacing Problem		
												A072202	Low Impedance		
												A090801	Signal Artifact/Noise		
												A1301	Failure to Read Input Signal		
										E2403	No Clinical Signs, Symptoms or Conditions	A070909	Oversensing		
												A0712	Pacing Problem		
												A072202	Low Impedance		
												A090801	Signal Artifact/Noise		
												A1301	Failure to Read Input Signal		
										F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A070909	Oversensing		
												A0712	Pacing Problem		
												A072202	Low Impedance		
												A090801	Signal Artifact/Noise		
												A1301	Failure to Read Input Signal		
										F15	Recognized Device or Procedural Complication	A070909	Oversensing		
												A0712	Pacing Problem		
												A072202	Low Impedance		
												A090801	Signal Artifact/Noise		
												A1301	Failure to Read Input Signal		
										F2203	Imaging Required	A070909	Oversensing		
												A0712	Pacing Problem		
												A072202	Low Impedance		
												A090801	Signal Artifact/Noise		
												A1301	Failure to Read Input Signal		
								CT2407	Active implantable	E060104	Bradycardia	A070909	Oversensing		
												A0712	Pacing Problem		
												A072202	Low Impedance		
												A090801	Signal Artifact/Noise		
												A1301	Failure to Read Input Signal		
										E2403	No Clinical Signs, Symptoms or Conditions	A070909	Oversensing		
												A0712	Pacing Problem		
												A072202	Low Impedance		
												A090801	Signal Artifact/Noise		
												A1301	Failure to Read Input Signal		
										F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A070909	Oversensing		
												A0712	Pacing Problem		
												A072202	Low Impedance		
												A090801	Signal Artifact/Noise		
												A1301	Failure to Read Input Signal		
										F15	Recognized Device or Procedural Complication	A070909	Oversensing		
												A0712	Pacing Problem		
												A072202	Low Impedance		
												A090801	Signal Artifact/Noise		
												A1301	Failure to Read Input Signal		

							F2203	Imaging Required	A070909	Oversensing		
									A0712	Pacing Problem		
									A072202	Low Impedance		
									A090801	Signal Artifact/Noise		
									A1301	Failure to Read Input Signal		
					CT257	Implantable cardiac pacemakers	E060104	Bradycardia	A070909	Oversensing		
									A0712	Pacing Problem		
									A072202	Low Impedance		
									A090801	Signal Artifact/Noise		
									A1301	Failure to Read Input Signal		
							E2403	No Clinical Signs, Symptoms or Conditions	A070909	Oversensing		
									A0712	Pacing Problem		
									A072202	Low Impedance		
									A090801	Signal Artifact/Noise		
									A1301	Failure to Read Input Signal		
							F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A070909	Oversensing		
									A0712	Pacing Problem		
									A072202	Low Impedance		
									A090801	Signal Artifact/Noise		
									A1301	Failure to Read Input Signal		
							F15	Recognized Device or Procedural Complication	A070909	Oversensing		
									A0712	Pacing Problem		
									A072202	Low Impedance		
									A090801	Signal Artifact/Noise		
									A1301	Failure to Read Input Signal		
							F2203	Imaging Required	A070909	Oversensing		
									A0712	Pacing Problem		
									A072202	Low Impedance		
									A090801	Signal Artifact/Noise		
									A1301	Failure to Read Input Signal		
					CT2727	Active implant devices/ancillaries	E060104	Bradycardia	A070909	Oversensing		
									A0712	Pacing Problem		
									A072202	Low Impedance		
									A090801	Signal Artifact/Noise		
									A1301	Failure to Read Input Signal		
							E2403	No Clinical Signs, Symptoms or Conditions	A070909	Oversensing		
									A0712	Pacing Problem		
									A072202	Low Impedance		
									A090801	Signal Artifact/Noise		
									A1301	Failure to Read Input Signal		
							F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A070909	Oversensing		
									A0712	Pacing Problem		
									A072202	Low Impedance		
									A090801	Signal Artifact/Noise		
									A1301	Failure to Read Input Signal		
							F15	Recognized Device or Procedural Complication	A070909	Oversensing		
									A0712	Pacing Problem		
									A072202	Low Impedance		
									A090801	Signal Artifact/Noise		
									A1301	Failure to Read Input Signal		
							F2203	Imaging Required	A070909	Oversensing		
									A0712	Pacing Problem		
									A072202	Low Impedance		
									A090801	Signal Artifact/Noise		
									A1301	Failure to Read Input Signal		
					CT2750	Implantable cardiac electrical stimulators and associated devices	E060104	Bradycardia	A070909	Oversensing		
									A0712	Pacing Problem		
									A072202	Low Impedance		
									A090801	Signal Artifact/Noise		
									A1301	Failure to Read Input Signal		
							E2403	No Clinical Signs, Symptoms or Conditions	A070909	Oversensing		
									A0712	Pacing Problem		
									A072202	Low Impedance		
									A090801	Signal Artifact/Noise		
									A1301	Failure to Read Input Signal		
							F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A070909	Oversensing		
									A0712	Pacing Problem		
									A072202	Low Impedance		
									A090801	Signal Artifact/Noise		
									A1301	Failure to Read Input Signal		
							F15	Recognized Device or Procedural Complication	A070909	Oversensing		
									A0712	Pacing Problem		
									A072202	Low Impedance		
									A090801	Signal Artifact/Noise		
									A1301	Failure to Read Input Signal		
							F2203	Imaging Required	A070909	Oversensing		
									A0712	Pacing Problem		
									A072202	Low Impedance		
									A090801	Signal Artifact/Noise		
									A1301	Failure to Read Input Signal		
					CT2866	Implantable cardiac electrical stimulators	E060104	Bradycardia	A070909	Oversensing		
									A0712	Pacing Problem		
									A072202	Low Impedance		
									A090801	Signal Artifact/Noise		
									A1301	Failure to Read Input Signal		
							E2403	No Clinical Signs, Symptoms or Conditions	A070909	Oversensing		
									A0712	Pacing Problem		
									A072202	Low Impedance		

									A090801	Signal Artifact/Noise		
									A1301	Failure to Read Input Signal		
							F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A070909	Oversensing		
									A0712	Pacing Problem		
									A072202	Low Impedance		
									A090801	Signal Artifact/Noise		
									A1301	Failure to Read Input Signal		
							F15	Recognized Device or Procedural Complication	A070909	Oversensing		
									A0712	Pacing Problem		
									A072202	Low Impedance		
									A090801	Signal Artifact/Noise		
									A1301	Failure to Read Input Signal		
							F2203	Imaging Required	A070909	Oversensing		
									A0712	Pacing Problem		
									A072202	Low Impedance		
									A090801	Signal Artifact/Noise		
									A1301	Failure to Read Input Signal		
						CT2985	Active	E060104	Bradycardia	A070909	Oversensing	
										A0712	Pacing Problem	
										A072202	Low Impedance	
										A090801	Signal Artifact/Noise	
										A1301	Failure to Read Input Signal	
								E2403	No Clinical Signs, Symptoms or Conditions	A070909	Oversensing	
										A0712	Pacing Problem	
										A072202	Low Impedance	
										A090801	Signal Artifact/Noise	
										A1301	Failure to Read Input Signal	
								F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A070909	Oversensing	
										A0712	Pacing Problem	
										A072202	Low Impedance	
										A090801	Signal Artifact/Noise	
										A1301	Failure to Read Input Signal	
								F15	Recognized Device or Procedural Complication	A070909	Oversensing	
										A0712	Pacing Problem	
										A072202	Low Impedance	
										A090801	Signal Artifact/Noise	
										A1301	Failure to Read Input Signal	
								F2203	Imaging Required	A070909	Oversensing	
										A0712	Pacing Problem	
										A072202	Low Impedance	
										A090801	Signal Artifact/Noise	
										A1301	Failure to Read Input Signal	
						CT3001	Implantable metal	E060104	Bradycardia	A070909	Oversensing	
										A0712	Pacing Problem	
										A072202	Low Impedance	
										A090801	Signal Artifact/Noise	
										A1301	Failure to Read Input Signal	
								E2403	No Clinical Signs, Symptoms or Conditions	A070909	Oversensing	
										A0712	Pacing Problem	
										A072202	Low Impedance	
										A090801	Signal Artifact/Noise	
										A1301	Failure to Read Input Signal	
								F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A070909	Oversensing	
										A0712	Pacing Problem	
										A072202	Low Impedance	
										A090801	Signal Artifact/Noise	
										A1301	Failure to Read Input Signal	
								F15	Recognized Device or Procedural Complication	A070909	Oversensing	
										A0712	Pacing Problem	
										A072202	Low Impedance	
										A090801	Signal Artifact/Noise	
										A1301	Failure to Read Input Signal	
								F2203	Imaging Required	A070909	Oversensing	
										A0712	Pacing Problem	
										A072202	Low Impedance	
										A090801	Signal Artifact/Noise	
										A1301	Failure to Read Input Signal	
						CT3002	Implantable synthetic polymer	E060104	Bradycardia	A070909	Oversensing	
										A0712	Pacing Problem	
										A072202	Low Impedance	
										A090801	Signal Artifact/Noise	
										A1301	Failure to Read Input Signal	
								E2403	No Clinical Signs, Symptoms or Conditions	A070909	Oversensing	
										A0712	Pacing Problem	
										A072202	Low Impedance	
										A090801	Signal Artifact/Noise	
										A1301	Failure to Read Input Signal	
								F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A070909	Oversensing	
										A0712	Pacing Problem	
										A072202	Low Impedance	
										A090801	Signal Artifact/Noise	
										A1301	Failure to Read Input Signal	
								F15	Recognized Device or Procedural Complication	A070909	Oversensing	
										A0712	Pacing Problem	
										A072202	Low Impedance	
										A090801	Signal Artifact/Noise	
										A1301	Failure to Read Input Signal	
								F2203	Imaging Required	A070909	Oversensing	

									A0712	Pacing Problem		
									A072202	Low Impedance		
									A090801	Signal Artifact/Noise		
									A1301	Failure to Read Input Signal		
					CT3006	Implantable inorganic material	E060104	Bradycardia	A070909	Oversensing		
									A0712	Pacing Problem		
									A072202	Low Impedance		
									A090801	Signal Artifact/Noise		
									A1301	Failure to Read Input Signal		
							E2403	No Clinical Signs, Symptoms or Conditions	A070909	Oversensing		
									A0712	Pacing Problem		
									A072202	Low Impedance		
									A090801	Signal Artifact/Noise		
									A1301	Failure to Read Input Signal		
							F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A070909	Oversensing		
									A0712	Pacing Problem		
									A072202	Low Impedance		
									A090801	Signal Artifact/Noise		
									A1301	Failure to Read Input Signal		
							F15	Recognized Device or Procedural Complication	A070909	Oversensing		
									A0712	Pacing Problem		
									A072202	Low Impedance		
									A090801	Signal Artifact/Noise		
									A1301	Failure to Read Input Signal		
							F2203	Imaging Required	A070909	Oversensing		
									A0712	Pacing Problem		
									A072202	Low Impedance		
									A090801	Signal Artifact/Noise		
									A1301	Failure to Read Input Signal		
					CT314	Electrically powered	E060104	Bradycardia	A070909	Oversensing		
									A0712	Pacing Problem		
									A072202	Low Impedance		
									A090801	Signal Artifact/Noise		
									A1301	Failure to Read Input Signal		
							E2403	No Clinical Signs, Symptoms or Conditions	A070909	Oversensing		
									A0712	Pacing Problem		
									A072202	Low Impedance		
									A090801	Signal Artifact/Noise		
									A1301	Failure to Read Input Signal		
							F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A070909	Oversensing		
									A0712	Pacing Problem		
									A072202	Low Impedance		
									A090801	Signal Artifact/Noise		
									A1301	Failure to Read Input Signal		
							F15	Recognized Device or Procedural Complication	A070909	Oversensing		
									A0712	Pacing Problem		
									A072202	Low Impedance		
									A090801	Signal Artifact/Noise		
									A1301	Failure to Read Input Signal		
							F2203	Imaging Required	A070909	Oversensing		
									A0712	Pacing Problem		
									A072202	Low Impedance		
									A090801	Signal Artifact/Noise		
									A1301	Failure to Read Input Signal		
					CT321	Long-term surgical invasive	E060104	Bradycardia	A070909	Oversensing		
									A0712	Pacing Problem		
									A072202	Low Impedance		
									A090801	Signal Artifact/Noise		
									A1301	Failure to Read Input Signal		
							E2403	No Clinical Signs, Symptoms or Conditions	A070909	Oversensing		
									A0712	Pacing Problem		
									A072202	Low Impedance		
									A090801	Signal Artifact/Noise		
									A1301	Failure to Read Input Signal		
							F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A070909	Oversensing		
									A0712	Pacing Problem		
									A072202	Low Impedance		
									A090801	Signal Artifact/Noise		
									A1301	Failure to Read Input Signal		
							F15	Recognized Device or Procedural Complication	A070909	Oversensing		
									A0712	Pacing Problem		
									A072202	Low Impedance		
									A090801	Signal Artifact/Noise		
									A1301	Failure to Read Input Signal		
							F2203	Imaging Required	A070909	Oversensing		
									A0712	Pacing Problem		
									A072202	Low Impedance		
									A090801	Signal Artifact/Noise		
									A1301	Failure to Read Input Signal		
					CT449	Generators and associated devices	E060104	Bradycardia	A070909	Oversensing		
									A0712	Pacing Problem		
									A072202	Low Impedance		
									A090801	Signal Artifact/Noise		
									A1301	Failure to Read Input Signal		
							E2403	No Clinical Signs, Symptoms or Conditions	A070909	Oversensing		
									A0712	Pacing Problem		
									A072202	Low Impedance		
									A090801	Signal Artifact/Noise		

									A1301	Failure to Read Input Signal		
							F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A070909	Oversensing		
									A0712	Pacing Problem		
									A072202	Low Impedance		
									A090801	Signal Artifact/Noise		
									A1301	Failure to Read Input Signal		
							F15	Recognized Device or Procedural Complication	A070909	Oversensing		
									A0712	Pacing Problem		
									A072202	Low Impedance		
									A090801	Signal Artifact/Noise		
									A1301	Failure to Read Input Signal		
							F2203	Imaging Required	A070909	Oversensing		
									A0712	Pacing Problem		
									A072202	Low Impedance		
									A090801	Signal Artifact/Noise		
									A1301	Failure to Read Input Signal		
					CT981	Single use	E060104	Bradycardia	A070909	Oversensing		
									A0712	Pacing Problem		
									A072202	Low Impedance		
									A090801	Signal Artifact/Noise		
									A1301	Failure to Read Input Signal		
							E2403	No Clinical Signs, Symptoms or Conditions	A070909	Oversensing		
									A0712	Pacing Problem		
									A072202	Low Impedance		
									A090801	Signal Artifact/Noise		
									A1301	Failure to Read Input Signal		
							F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A070909	Oversensing		
									A0712	Pacing Problem		
									A072202	Low Impedance		
									A090801	Signal Artifact/Noise		
									A1301	Failure to Read Input Signal		
							F15	Recognized Device or Procedural Complication	A070909	Oversensing		
									A0712	Pacing Problem		
									A072202	Low Impedance		
									A090801	Signal Artifact/Noise		
									A1301	Failure to Read Input Signal		
							F2203	Imaging Required	A070909	Oversensing		
									A0712	Pacing Problem		
									A072202	Low Impedance		
									A090801	Signal Artifact/Noise		
									A1301	Failure to Read Input Signal		
					CT983	Surgical invasive	E060104	Bradycardia	A070909	Oversensing		
									A0712	Pacing Problem		
									A072202	Low Impedance		
									A090801	Signal Artifact/Noise		
									A1301	Failure to Read Input Signal		
							E2403	No Clinical Signs, Symptoms or Conditions	A070909	Oversensing		
									A0712	Pacing Problem		
									A072202	Low Impedance		
									A090801	Signal Artifact/Noise		
									A1301	Failure to Read Input Signal		
							F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A070909	Oversensing		
									A0712	Pacing Problem		
									A072202	Low Impedance		
									A090801	Signal Artifact/Noise		
									A1301	Failure to Read Input Signal		
							F15	Recognized Device or Procedural Complication	A070909	Oversensing		
									A0712	Pacing Problem		
									A072202	Low Impedance		
									A090801	Signal Artifact/Noise		
									A1301	Failure to Read Input Signal		
							F2203	Imaging Required	A070909	Oversensing		
									A0712	Pacing Problem		
									A072202	Low Impedance		
									A090801	Signal Artifact/Noise		
									A1301	Failure to Read Input Signal		
6615	2021	All other reportable incidents	36241	Pacemaker/defibrillator lead	CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A0203	Defective Device	19/01/2022	19/01/2022
									A040101	Fracture	19/01/2022	19/01/2022
									A070102	High Capture Threshold	19/01/2022	19/01/2022
									A070909	Oversensing	19/01/2022	19/01/2022
									A072202	Low Impedance	19/01/2022	19/01/2022
									A090801	Signal Artifact/Noise	19/01/2022	19/01/2022
							F08	Hospitalization or Prolonged Hospitalization	A0203	Defective Device	19/01/2022	19/01/2022
									A040101	Fracture	19/01/2022	19/01/2022
									A070102	High Capture Threshold	19/01/2022	19/01/2022
									A070909	Oversensing	19/01/2022	19/01/2022
									A072202	Low Impedance	19/01/2022	19/01/2022
									A090801	Signal Artifact/Noise	19/01/2022	19/01/2022
							F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A0203	Defective Device	19/01/2022	19/01/2022
									A040101	Fracture	19/01/2022	19/01/2022
									A070102	High Capture Threshold	19/01/2022	19/01/2022
									A070909	Oversensing	19/01/2022	19/01/2022
									A072202	Low Impedance	19/01/2022	19/01/2022
									A090801	Signal Artifact/Noise	19/01/2022	19/01/2022
							F1905	Device Revision or Replacement	A0203	Defective Device	19/01/2022	19/01/2022
									A040101	Fracture	19/01/2022	19/01/2022
									A070102	High Capture Threshold	19/01/2022	19/01/2022
									A070909	Oversensing	19/01/2022	19/01/2022

									A072202	Low Impedance	19/01/2022	19/01/2022
									A090801	Signal Artifact/Noise	19/01/2022	19/01/2022
							F2203	Imaging Required	A0203	Defective Device	19/01/2022	19/01/2022
									A040101	Fracture	19/01/2022	19/01/2022
									A070102	High Capture Threshold	19/01/2022	19/01/2022
									A070909	Oversensing	19/01/2022	19/01/2022
									A072202	Low Impedance	19/01/2022	19/01/2022
									A090801	Signal Artifact/Noise	19/01/2022	19/01/2022
6616	2022	All other reportable incidents	17882	Physiologic-monitoring defibrillation system	CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A0719	Unexpected Shutdown		
							F26	No Health Consequences or Impact	A0719	Unexpected Shutdown		
6617	2021	All other reportable incidents	35223	Endocardial/interventricular septal pacing lead	CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A0401	Break	16/09/2021	16/09/2021
									A070909	Oversensing	16/09/2021	16/09/2021
							F1901	Additional Surgery	A0401	Break	16/09/2021	16/09/2021
									A070909	Oversensing	16/09/2021	16/09/2021
6618	2021	All other reportable incidents	36241	Pacemaker/defibrillator lead	CT1684	Cardiac pacemakers and associated devices	E2104	Electric Shock	A071302	Inappropriate/Inadequate Shock/Stimulation	30/03/2022	30/03/2022
									A072202	Low Impedance	30/03/2022	30/03/2022
							E2403	No Clinical Signs, Symptoms or Conditions	A071302	Inappropriate/Inadequate Shock/Stimulation	30/03/2022	30/03/2022
									A072202	Low Impedance	30/03/2022	30/03/2022
							F08	Hospitalization or Prolonged Hospitalization	A071302	Inappropriate/Inadequate Shock/Stimulation	30/03/2022	30/03/2022
									A072202	Low Impedance	30/03/2022	30/03/2022
							F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A071302	Inappropriate/Inadequate Shock/Stimulation	30/03/2022	30/03/2022
									A072202	Low Impedance	30/03/2022	30/03/2022
							F1905	Device Revision or Replacement	A071302	Inappropriate/Inadequate Shock/Stimulation	30/03/2022	30/03/2022
									A072202	Low Impedance	30/03/2022	30/03/2022
6619	2021	All other reportable incidents	47267	Single-chamber implantable pacemaker, rate-responsive	CT141	Electrodes/Leads and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	16/09/2021	16/09/2021
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	16/09/2021	16/09/2021
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	16/09/2021	16/09/2021
					CT145	Cardiovascular devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	16/09/2021	16/09/2021
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	16/09/2021	16/09/2021
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	16/09/2021	16/09/2021
					CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	16/09/2021	16/09/2021
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	16/09/2021	16/09/2021
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	16/09/2021	16/09/2021
					CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	16/09/2021	16/09/2021
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	16/09/2021	16/09/2021
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	16/09/2021	16/09/2021
					CT233	Surgical	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	16/09/2021	16/09/2021
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	16/09/2021	16/09/2021
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	16/09/2021	16/09/2021
					CT2406	Implantable	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	16/09/2021	16/09/2021
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	16/09/2021	16/09/2021
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	16/09/2021	16/09/2021
					CT2407	Active implantable	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	16/09/2021	16/09/2021
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	16/09/2021	16/09/2021
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	16/09/2021	16/09/2021
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	16/09/2021	16/09/2021
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	16/09/2021	16/09/2021
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	16/09/2021	16/09/2021
					CT2727	Active implant devices/ancillaries	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	16/09/2021	16/09/2021
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	16/09/2021	16/09/2021
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	16/09/2021	16/09/2021
					CT2750	Implantable cardiac electrical stimulators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	16/09/2021	16/09/2021
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	16/09/2021	16/09/2021
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	16/09/2021	16/09/2021
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	16/09/2021	16/09/2021
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	16/09/2021	16/09/2021
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	16/09/2021	16/09/2021
					CT2985	Active	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	16/09/2021	16/09/2021
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	16/09/2021	16/09/2021
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	16/09/2021	16/09/2021
					CT3001	Implantable metal	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	16/09/2021	16/09/2021
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	16/09/2021	16/09/2021
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	16/09/2021	16/09/2021
					CT3002	Implantable synthetic polymer	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	16/09/2021	16/09/2021
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	16/09/2021	16/09/2021
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	16/09/2021	16/09/2021
					CT3006	Implantable inorganic material	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	16/09/2021	16/09/2021
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	16/09/2021	16/09/2021
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	16/09/2021	16/09/2021
					CT314	Electrically powered	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	16/09/2021	16/09/2021
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	16/09/2021	16/09/2021
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	16/09/2021	16/09/2021
					CT321	Long-term surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	16/09/2021	16/09/2021
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	16/09/2021	16/09/2021
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	16/09/2021	16/09/2021
					CT449	Generators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	16/09/2021	16/09/2021
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	16/09/2021	16/09/2021
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	16/09/2021	16/09/2021
					CT981	Single use	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	16/09/2021	16/09/2021
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	16/09/2021	16/09/2021
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	16/09/2021	16/09/2021
					CT983	Surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	16/09/2021	16/09/2021
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	16/09/2021	16/09/2021
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	16/09/2021	16/09/2021
6620	2021	All other reportable incidents	47263	Cardiac resynchronization therapy implantable pacemaker	CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A0908	Incorrect, Inadequate or Imprecise Result or Readings	19/08/2021	19/08/2021
							F08	Hospitalization or Prolonged Hospitalization	A0908	Incorrect, Inadequate or Imprecise Result or Readings	19/08/2021	19/08/2021

						F1903	Device Explantation	A0908	Incorrect, Inadequate or Imprecise Result or Readings	19/08/2021	19/08/2021
						F1905	Device Revision or Replacement	A0908	Incorrect, Inadequate or Imprecise Result or Readings	19/08/2021	19/08/2021
					CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A0908	Incorrect, Inadequate or Imprecise Result or Readings	19/08/2021	19/08/2021
							F08	Hospitalization or Prolonged Hospitalization	A0908	Incorrect, Inadequate or Imprecise Result or Readings	19/08/2021	19/08/2021
							F1903	Device Explantation	A0908	Incorrect, Inadequate or Imprecise Result or Readings	19/08/2021	19/08/2021
							F1905	Device Revision or Replacement	A0908	Incorrect, Inadequate or Imprecise Result or Readings	19/08/2021	19/08/2021
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A0908	Incorrect, Inadequate or Imprecise Result or Readings	19/08/2021	19/08/2021
							F08	Hospitalization or Prolonged Hospitalization	A0908	Incorrect, Inadequate or Imprecise Result or Readings	19/08/2021	19/08/2021
							F1903	Device Explantation	A0908	Incorrect, Inadequate or Imprecise Result or Readings	19/08/2021	19/08/2021
							F1905	Device Revision or Replacement	A0908	Incorrect, Inadequate or Imprecise Result or Readings	19/08/2021	19/08/2021
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A0908	Incorrect, Inadequate or Imprecise Result or Readings	19/08/2021	19/08/2021
							F08	Hospitalization or Prolonged Hospitalization	A0908	Incorrect, Inadequate or Imprecise Result or Readings	19/08/2021	19/08/2021
							F1903	Device Explantation	A0908	Incorrect, Inadequate or Imprecise Result or Readings	19/08/2021	19/08/2021
							F1905	Device Revision or Replacement	A0908	Incorrect, Inadequate or Imprecise Result or Readings	19/08/2021	19/08/2021
6621	2021	All other reportable incidents	35852	Single-chamber implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem	27/10/2021	27/10/2021
									A070504	Premature Discharge of Battery	27/10/2021	27/10/2021
							F08	Hospitalization or Prolonged Hospitalization	A0705	Battery Problem	27/10/2021	27/10/2021
									A070504	Premature Discharge of Battery	27/10/2021	27/10/2021
							F1905	Device Revision or Replacement	A0705	Battery Problem	27/10/2021	27/10/2021
									A070504	Premature Discharge of Battery	27/10/2021	27/10/2021
6622	2021	All other reportable incidents	35852	Single-chamber implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem	21/10/2021	21/10/2021
									A070504	Premature Discharge of Battery	21/10/2021	21/10/2021
							F08	Hospitalization or Prolonged Hospitalization	A0705	Battery Problem	21/10/2021	21/10/2021
									A070504	Premature Discharge of Battery	21/10/2021	21/10/2021
							F1905	Device Revision or Replacement	A0705	Battery Problem	21/10/2021	21/10/2021
									A070504	Premature Discharge of Battery	21/10/2021	21/10/2021
6623	2021	All other reportable incidents	47263	Cardiac resynchronization therapy implantable pacemaker	CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A0908	Incorrect, Inadequate or Imprecise Result or Readings	09/09/2021	09/09/2021
							F08	Hospitalization or Prolonged Hospitalization	A0908	Incorrect, Inadequate or Imprecise Result or Readings	09/09/2021	09/09/2021
							F1905	Device Revision or Replacement	A0908	Incorrect, Inadequate or Imprecise Result or Readings	09/09/2021	09/09/2021
					CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A0908	Incorrect, Inadequate or Imprecise Result or Readings	09/09/2021	09/09/2021
							F08	Hospitalization or Prolonged Hospitalization	A0908	Incorrect, Inadequate or Imprecise Result or Readings	09/09/2021	09/09/2021
							F1905	Device Revision or Replacement	A0908	Incorrect, Inadequate or Imprecise Result or Readings	09/09/2021	09/09/2021
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A0908	Incorrect, Inadequate or Imprecise Result or Readings	09/09/2021	09/09/2021
							F08	Hospitalization or Prolonged Hospitalization	A0908	Incorrect, Inadequate or Imprecise Result or Readings	09/09/2021	09/09/2021
							F1905	Device Revision or Replacement	A0908	Incorrect, Inadequate or Imprecise Result or Readings	09/09/2021	09/09/2021
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A0908	Incorrect, Inadequate or Imprecise Result or Readings	09/09/2021	09/09/2021
							F08	Hospitalization or Prolonged Hospitalization	A0908	Incorrect, Inadequate or Imprecise Result or Readings	09/09/2021	09/09/2021
							F1905	Device Revision or Replacement	A0908	Incorrect, Inadequate or Imprecise Result or Readings	09/09/2021	09/09/2021
6624	2022	All other reportable incidents	47267	Single-chamber implantable pacemaker, rate-responsive	CT141	Electrodes/Leads and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A071102	Failure to Interrogate		
							F24	Insufficient Health Impact Information	A071102	Failure to Interrogate		
					CT145	Cardiovascular devices	E2403	No Clinical Signs, Symptoms or Conditions	A071102	Failure to Interrogate		
							F24	Insufficient Health Impact Information	A071102	Failure to Interrogate		
					CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A071102	Failure to Interrogate		
							F24	Insufficient Health Impact Information	A071102	Failure to Interrogate		
					CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A071102	Failure to Interrogate		
							F24	Insufficient Health Impact Information	A071102	Failure to Interrogate		
					CT233	Surgical	E2403	No Clinical Signs, Symptoms or Conditions	A071102	Failure to Interrogate		
							F24	Insufficient Health Impact Information	A071102	Failure to Interrogate		
					CT2406	Implantable	E2403	No Clinical Signs, Symptoms or Conditions	A071102	Failure to Interrogate		
							F24	Insufficient Health Impact Information	A071102	Failure to Interrogate		
					CT2407	Active implantable	E2403	No Clinical Signs, Symptoms or Conditions	A071102	Failure to Interrogate		
							F24	Insufficient Health Impact Information	A071102	Failure to Interrogate		
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A071102	Failure to Interrogate		
							F24	Insufficient Health Impact Information	A071102	Failure to Interrogate		
					CT2727	Active implant devices/ancillaries	E2403	No Clinical Signs, Symptoms or Conditions	A071102	Failure to Interrogate		
							F24	Insufficient Health Impact Information	A071102	Failure to Interrogate		
					CT2750	Implantable cardiac electrical stimulators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A071102	Failure to Interrogate		
							F24	Insufficient Health Impact Information	A071102	Failure to Interrogate		
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A071102	Failure to Interrogate		
							F24	Insufficient Health Impact Information	A071102	Failure to Interrogate		
					CT2985	Active	E2403	No Clinical Signs, Symptoms or Conditions	A071102	Failure to Interrogate		
							F24	Insufficient Health Impact Information	A071102	Failure to Interrogate		
					CT3001	Implantable metal	E2403	No Clinical Signs, Symptoms or Conditions	A071102	Failure to Interrogate		
							F24	Insufficient Health Impact Information	A071102	Failure to Interrogate		
					CT3002	Implantable synthetic polymer	E2403	No Clinical Signs, Symptoms or Conditions	A071102	Failure to Interrogate		
							F24	Insufficient Health Impact Information	A071102	Failure to Interrogate		
					CT3006	Implantable inorganic material	E2403	No Clinical Signs, Symptoms or Conditions	A071102	Failure to Interrogate		
							F24	Insufficient Health Impact Information	A071102	Failure to Interrogate		

					CT314	Electrically powered	E2403	No Clinical Signs, Symptoms or Conditions	A071102	Failure to Interrogate		
							F24	Insufficient Health Impact Information	A071102	Failure to Interrogate		
					CT321	Long-term surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A071102	Failure to Interrogate		
							F24	Insufficient Health Impact Information	A071102	Failure to Interrogate		
					CT449	Generators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A071102	Failure to Interrogate		
							F24	Insufficient Health Impact Information	A071102	Failure to Interrogate		
					CT981	Single use	E2403	No Clinical Signs, Symptoms or Conditions	A071102	Failure to Interrogate		
							F24	Insufficient Health Impact Information	A071102	Failure to Interrogate		
					CT983	Surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A071102	Failure to Interrogate		
							F24	Insufficient Health Impact Information	A071102	Failure to Interrogate		
6625	2021	All other reportable incidents	35822	External pacemaker, transcutaneous pacing	CT1684	Cardiac pacemakers and associated devices	E2401	Insufficient Information	A070101	Failure to Capture		
							E2403	No Clinical Signs, Symptoms or Conditions	A070101	Failure to Capture		
							F24	Insufficient Health Impact Information	A070101	Failure to Capture		
					CT1685	Cardiac pacemakers	E2401	Insufficient Information	A070101	Failure to Capture		
							E2403	No Clinical Signs, Symptoms or Conditions	A070101	Failure to Capture		
							F24	Insufficient Health Impact Information	A070101	Failure to Capture		
6626	2022	All other reportable incidents	47265	Dual-chamber implantable pacemaker, rate-responsive	CT141	Electrodes/Leads and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	13/04/2022	13/04/2022
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	13/04/2022	13/04/2022
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	13/04/2022	13/04/2022
					CT145	Cardiovascular devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	13/04/2022	13/04/2022
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	13/04/2022	13/04/2022
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	13/04/2022	13/04/2022
					CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	13/04/2022	13/04/2022
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	13/04/2022	13/04/2022
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	13/04/2022	13/04/2022
					CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	13/04/2022	13/04/2022
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	13/04/2022	13/04/2022
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	13/04/2022	13/04/2022
					CT233	Surgical	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	13/04/2022	13/04/2022
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	13/04/2022	13/04/2022
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	13/04/2022	13/04/2022
					CT2406	Implantable	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	13/04/2022	13/04/2022
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	13/04/2022	13/04/2022
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	13/04/2022	13/04/2022
					CT2407	Active implantable	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	13/04/2022	13/04/2022
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	13/04/2022	13/04/2022
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	13/04/2022	13/04/2022
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	13/04/2022	13/04/2022
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	13/04/2022	13/04/2022
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	13/04/2022	13/04/2022
					CT2727	Active implant devices/ancillaries	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	13/04/2022	13/04/2022
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	13/04/2022	13/04/2022
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	13/04/2022	13/04/2022
					CT2750	Implantable cardiac electrical stimulators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	13/04/2022	13/04/2022
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	13/04/2022	13/04/2022
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	13/04/2022	13/04/2022
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	13/04/2022	13/04/2022
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	13/04/2022	13/04/2022
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	13/04/2022	13/04/2022
					CT2985	Active	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	13/04/2022	13/04/2022
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	13/04/2022	13/04/2022
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	13/04/2022	13/04/2022
					CT3001	Implantable metal	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	13/04/2022	13/04/2022
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	13/04/2022	13/04/2022
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	13/04/2022	13/04/2022
					CT3002	Implantable synthetic polymer	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	13/04/2022	13/04/2022
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	13/04/2022	13/04/2022
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	13/04/2022	13/04/2022
					CT3006	Implantable inorganic material	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	13/04/2022	13/04/2022
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	13/04/2022	13/04/2022
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	13/04/2022	13/04/2022
					CT314	Electrically powered	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	13/04/2022	13/04/2022
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	13/04/2022	13/04/2022
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	13/04/2022	13/04/2022
					CT321	Long-term surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	13/04/2022	13/04/2022
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	13/04/2022	13/04/2022
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	13/04/2022	13/04/2022
					CT449	Generators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	13/04/2022	13/04/2022
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	13/04/2022	13/04/2022
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	13/04/2022	13/04/2022
					CT981	Single use	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	13/04/2022	13/04/2022
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	13/04/2022	13/04/2022
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	13/04/2022	13/04/2022
					CT983	Surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	13/04/2022	13/04/2022
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	13/04/2022	13/04/2022
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	13/04/2022	13/04/2022
6627	2022	All other reportable incidents	47263	Cardiac resynchronization therapy implantable pacemaker	CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	18/05/2022	18/05/2022
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	18/05/2022	18/05/2022
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	18/05/2022	18/05/2022
					CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	18/05/2022	18/05/2022
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	18/05/2022	18/05/2022
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	18/05/2022	18/05/2022
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	18/05/2022	18/05/2022
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	18/05/2022	18/05/2022
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	18/05/2022	18/05/2022
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	18/05/2022	18/05/2022
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	18/05/2022	18/05/2022
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	18/05/2022	18/05/2022

6628	2022	All other reportable incidents	37265	Dual-chamber implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A072201	High Impedance		
							F26	No Health Consequences or Impact	A072201	High Impedance		
6629	2022	All other reportable incidents	36241	Pacemaker/defibrillator lead	CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A072201	High Impedance		
							F26	No Health Consequences or Impact	A072201	High Impedance		
6630	2022	All other reportable incidents	35223	Endocardial/interventricular septal pacing lead	CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070909	Oversensing		
									A0722	Impedance Problem		
							F26	No Health Consequences or Impact	A070909	Oversensing		
									A0722	Impedance Problem		
6631	2022	All other reportable incidents	35852	Single-chamber implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
6632	2022	All other reportable incidents	35852	Single-chamber implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem	31/08/2022	31/08/2022
									A070504	Premature Discharge of Battery	31/08/2022	31/08/2022
							F08	Hospitalization or Prolonged Hospitalization	A0705	Battery Problem	31/08/2022	31/08/2022
									A070504	Premature Discharge of Battery	31/08/2022	31/08/2022
							F1905	Device Revision or Replacement	A0705	Battery Problem	31/08/2022	31/08/2022
									A070504	Premature Discharge of Battery	31/08/2022	31/08/2022
6633	2021	All other reportable incidents	47265	Dual-chamber implantable pacemaker, rate-responsive	CT141	Electrodes/Leads and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	24/09/2021	24/09/2021
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	24/09/2021	24/09/2021
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	24/09/2021	24/09/2021
					CT145	Cardiovascular devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	24/09/2021	24/09/2021
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	24/09/2021	24/09/2021
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	24/09/2021	24/09/2021
					CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	24/09/2021	24/09/2021
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	24/09/2021	24/09/2021
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	24/09/2021	24/09/2021
					CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	24/09/2021	24/09/2021
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	24/09/2021	24/09/2021
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	24/09/2021	24/09/2021
					CT233	Surgical	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	24/09/2021	24/09/2021
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	24/09/2021	24/09/2021
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	24/09/2021	24/09/2021
					CT2406	Implantable	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	24/09/2021	24/09/2021
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	24/09/2021	24/09/2021
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	24/09/2021	24/09/2021
					CT2407	Active implantable	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	24/09/2021	24/09/2021
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	24/09/2021	24/09/2021
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	24/09/2021	24/09/2021
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	24/09/2021	24/09/2021
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	24/09/2021	24/09/2021
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	24/09/2021	24/09/2021
					CT2727	Active implant devices/ancillaries	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	24/09/2021	24/09/2021
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	24/09/2021	24/09/2021
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	24/09/2021	24/09/2021
					CT2750	Implantable cardiac electrical stimulators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	24/09/2021	24/09/2021
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	24/09/2021	24/09/2021
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	24/09/2021	24/09/2021
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	24/09/2021	24/09/2021
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	24/09/2021	24/09/2021
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	24/09/2021	24/09/2021
					CT2985	Active	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	24/09/2021	24/09/2021
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	24/09/2021	24/09/2021
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	24/09/2021	24/09/2021
					CT3001	Implantable metal	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	24/09/2021	24/09/2021
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	24/09/2021	24/09/2021
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	24/09/2021	24/09/2021
					CT3002	Implantable synthetic polymer	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	24/09/2021	24/09/2021
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	24/09/2021	24/09/2021
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	24/09/2021	24/09/2021
					CT3006	Implantable inorganic material	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	24/09/2021	24/09/2021
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	24/09/2021	24/09/2021
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	24/09/2021	24/09/2021
					CT314	Electrically powered	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	24/09/2021	24/09/2021
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	24/09/2021	24/09/2021
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	24/09/2021	24/09/2021
					CT321	Long-term surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	24/09/2021	24/09/2021
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	24/09/2021	24/09/2021
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	24/09/2021	24/09/2021
					CT449	Generators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	24/09/2021	24/09/2021
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	24/09/2021	24/09/2021
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	24/09/2021	24/09/2021
					CT981	Single use	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	24/09/2021	24/09/2021
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	24/09/2021	24/09/2021
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	24/09/2021	24/09/2021
					CT983	Surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	24/09/2021	24/09/2021
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	24/09/2021	24/09/2021
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	24/09/2021	24/09/2021
6634	2021	All other reportable incidents	47265	Dual-chamber implantable pacemaker, rate-responsive	CT141	Electrodes/Leads and associated devices	E0119	Loss of Consciousness	A070504	Premature Discharge of Battery		
							E011903	Syncope/Fainting	A070504	Premature Discharge of Battery		
							E0611	Heart Failure/Congestive Heart Failure	A070504	Premature Discharge of Battery		
							E0717	Dyspnea	A070504	Premature Discharge of Battery		
							E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery		
					CT145	Cardiovascular devices	E0119	Loss of Consciousness	A070504	Premature Discharge of Battery		
							E011903	Syncope/Fainting	A070504	Premature Discharge of Battery		
							E0611	Heart Failure/Congestive Heart Failure	A070504	Premature Discharge of Battery		
							E0717	Dyspnea	A070504	Premature Discharge of Battery		
							E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery		

					CT1684	Cardiac pacemakers and associated devices	E0119	Loss of Consciousness	A070504	Premature Discharge of Battery		
							E011903	Syncope/Fainting	A070504	Premature Discharge of Battery		
							E0611	Heart Failure/Congestive Heart Failure	A070504	Premature Discharge of Battery		
							E0717	Dyspnea	A070504	Premature Discharge of Battery		
							E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery		
					CT1685	Cardiac pacemakers	E0119	Loss of Consciousness	A070504	Premature Discharge of Battery		
							E011903	Syncope/Fainting	A070504	Premature Discharge of Battery		
							E0611	Heart Failure/Congestive Heart Failure	A070504	Premature Discharge of Battery		
							E0717	Dyspnea	A070504	Premature Discharge of Battery		
							E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery		
					CT233	Surgical	E0119	Loss of Consciousness	A070504	Premature Discharge of Battery		
							E011903	Syncope/Fainting	A070504	Premature Discharge of Battery		
							E0611	Heart Failure/Congestive Heart Failure	A070504	Premature Discharge of Battery		
							E0717	Dyspnea	A070504	Premature Discharge of Battery		
							E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery		
					CT2406	Implantable	E0119	Loss of Consciousness	A070504	Premature Discharge of Battery		
							E011903	Syncope/Fainting	A070504	Premature Discharge of Battery		
							E0611	Heart Failure/Congestive Heart Failure	A070504	Premature Discharge of Battery		
							E0717	Dyspnea	A070504	Premature Discharge of Battery		
							E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery		
					CT2407	Active implantable	E0119	Loss of Consciousness	A070504	Premature Discharge of Battery		
							E011903	Syncope/Fainting	A070504	Premature Discharge of Battery		
							E0611	Heart Failure/Congestive Heart Failure	A070504	Premature Discharge of Battery		
							E0717	Dyspnea	A070504	Premature Discharge of Battery		
							E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery		
					CT257	Implantable cardiac pacemakers	E0119	Loss of Consciousness	A070504	Premature Discharge of Battery		
							E011903	Syncope/Fainting	A070504	Premature Discharge of Battery		
							E0611	Heart Failure/Congestive Heart Failure	A070504	Premature Discharge of Battery		
							E0717	Dyspnea	A070504	Premature Discharge of Battery		
							E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery		
					CT2727	Active implant devices/ancillaries	E0119	Loss of Consciousness	A070504	Premature Discharge of Battery		
							E011903	Syncope/Fainting	A070504	Premature Discharge of Battery		
							E0611	Heart Failure/Congestive Heart Failure	A070504	Premature Discharge of Battery		
							E0717	Dyspnea	A070504	Premature Discharge of Battery		
							E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery		
					CT2750	Implantable cardiac electrical stimulators and associated devices	E0119	Loss of Consciousness	A070504	Premature Discharge of Battery		
							E011903	Syncope/Fainting	A070504	Premature Discharge of Battery		
							E0611	Heart Failure/Congestive Heart Failure	A070504	Premature Discharge of Battery		
							E0717	Dyspnea	A070504	Premature Discharge of Battery		
							E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery		
					CT2866	Implantable cardiac electrical stimulators	E0119	Loss of Consciousness	A070504	Premature Discharge of Battery		
							E011903	Syncope/Fainting	A070504	Premature Discharge of Battery		
							E0611	Heart Failure/Congestive Heart Failure	A070504	Premature Discharge of Battery		
							E0717	Dyspnea	A070504	Premature Discharge of Battery		
							E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery		
					CT2985	Active	E0119	Loss of Consciousness	A070504	Premature Discharge of Battery		
							E011903	Syncope/Fainting	A070504	Premature Discharge of Battery		
							E0611	Heart Failure/Congestive Heart Failure	A070504	Premature Discharge of Battery		
							E0717	Dyspnea	A070504	Premature Discharge of Battery		
							E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery		
					CT3001	Implantable metal	E0119	Loss of Consciousness	A070504	Premature Discharge of Battery		
							E011903	Syncope/Fainting	A070504	Premature Discharge of Battery		
							E0611	Heart Failure/Congestive Heart Failure	A070504	Premature Discharge of Battery		
							E0717	Dyspnea	A070504	Premature Discharge of Battery		
							E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery		
					CT3002	Implantable synthetic polymer	E0119	Loss of Consciousness	A070504	Premature Discharge of Battery		
							E011903	Syncope/Fainting	A070504	Premature Discharge of Battery		
							E0611	Heart Failure/Congestive Heart Failure	A070504	Premature Discharge of Battery		
							E0717	Dyspnea	A070504	Premature Discharge of Battery		
							E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery		
					CT3006	Implantable inorganic material	E0119	Loss of Consciousness	A070504	Premature Discharge of Battery		
							E011903	Syncope/Fainting	A070504	Premature Discharge of Battery		
							E0611	Heart Failure/Congestive Heart Failure	A070504	Premature Discharge of Battery		
							E0717	Dyspnea	A070504	Premature Discharge of Battery		
							E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery		
					CT314	Electrically powered	E0119	Loss of Consciousness	A070504	Premature Discharge of Battery		
							E011903	Syncope/Fainting	A070504	Premature Discharge of Battery		
							E0611	Heart Failure/Congestive Heart Failure	A070504	Premature Discharge of Battery		
							E0717	Dyspnea	A070504	Premature Discharge of Battery		
							E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery		
					CT321	Long-term surgical invasive	E0119	Loss of Consciousness	A070504	Premature Discharge of Battery		
							E011903	Syncope/Fainting	A070504	Premature Discharge of Battery		
							E0611	Heart Failure/Congestive Heart Failure	A070504	Premature Discharge of Battery		
							E0717	Dyspnea	A070504	Premature Discharge of Battery		

						E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery			
						F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery			
					CT449	Generators and associated devices	E0119	Loss of Consciousness	A070504	Premature Discharge of Battery		
							E011903	Syncope/Fainting	A070504	Premature Discharge of Battery		
							E0611	Heart Failure/Congestive Heart Failure	A070504	Premature Discharge of Battery		
							E0717	Dyspnea	A070504	Premature Discharge of Battery		
							E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery		
					CT981	Single use	E0119	Loss of Consciousness	A070504	Premature Discharge of Battery		
							E011903	Syncope/Fainting	A070504	Premature Discharge of Battery		
							E0611	Heart Failure/Congestive Heart Failure	A070504	Premature Discharge of Battery		
							E0717	Dyspnea	A070504	Premature Discharge of Battery		
							E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery		
					CT983	Surgical invasive	E0119	Loss of Consciousness	A070504	Premature Discharge of Battery		
							E011903	Syncope/Fainting	A070504	Premature Discharge of Battery		
							E0611	Heart Failure/Congestive Heart Failure	A070504	Premature Discharge of Battery		
							E0717	Dyspnea	A070504	Premature Discharge of Battery		
							E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery		
6635	2021	All other reportable incidents	35852	Single-chamber implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	07/10/2021	07/10/2021
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	07/10/2021	07/10/2021
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	07/10/2021	07/10/2021
6636	2021	All other reportable incidents	47265	Dual-chamber implantable pacemaker, rate-responsive	CT141	Electrodes/Leads and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A070504	Premature Discharge of Battery		
					CT145	Cardiovascular devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A070504	Premature Discharge of Battery		
					CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A070504	Premature Discharge of Battery		
					CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A070504	Premature Discharge of Battery		
					CT233	Surgical	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A070504	Premature Discharge of Battery		
					CT2406	Implantable	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A070504	Premature Discharge of Battery		
					CT2407	Active implantable	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A070504	Premature Discharge of Battery		
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A070504	Premature Discharge of Battery		
					CT2727	Active implant devices/ancillaries	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A070504	Premature Discharge of Battery		
					CT2750	Implantable cardiac electrical stimulators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A070504	Premature Discharge of Battery		
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A070504	Premature Discharge of Battery		
					CT2985	Active	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A070504	Premature Discharge of Battery		
					CT3001	Implantable metal	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A070504	Premature Discharge of Battery		
					CT3002	Implantable synthetic polymer	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A070504	Premature Discharge of Battery		
					CT3006	Implantable inorganic material	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A070504	Premature Discharge of Battery		
					CT314	Electrically powered	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A070504	Premature Discharge of Battery		
					CT321	Long-term surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A070504	Premature Discharge of Battery		
					CT449	Generators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A070504	Premature Discharge of Battery		
					CT981	Single use	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A070504	Premature Discharge of Battery		
					CT983	Surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A070504	Premature Discharge of Battery		
6637	2021	All other reportable incidents	36241	Pacemaker/defibrillator lead	CT1684	Cardiac pacemakers and associated devices	E2103	Device Overstimulation of Tissue	A051201	Device Dislodged or Dislocated		
							E2403	No Clinical Signs, Symptoms or Conditions	A051201	Device Dislodged or Dislocated		
							F08	Hospitalization or Prolonged Hospitalization	A051201	Device Dislodged or Dislocated		
							F1904	Device Repositioning	A051201	Device Dislodged or Dislocated		
							F23	Unexpected Medical Intervention	A051201	Device Dislodged or Dislocated		
6638	2022	All other reportable incidents	36241	Pacemaker/defibrillator lead	CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070909	Oversensing		
									A0712	Pacing Problem		
									A090801	Signal Artifact/Noise		
							F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A070909	Oversensing		
									A0712	Pacing Problem		
									A090801	Signal Artifact/Noise		
6639	2022	All other reportable incidents	36241	Pacemaker/defibrillator lead	CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A051201	Device Dislodged or Dislocated	09/05/2022	09/05/2022
							F08	Hospitalization or Prolonged Hospitalization	A051201	Device Dislodged or Dislocated	09/05/2022	09/05/2022
							F1905	Device Revision or Replacement	A051201	Device Dislodged or Dislocated	09/05/2022	09/05/2022
6640	2022	All other reportable incidents	35852	Single-chamber implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem	19/07/2022	19/07/2022
									A070504	Premature Discharge of Battery	19/07/2022	19/07/2022
							F08	Hospitalization or Prolonged Hospitalization	A0705	Battery Problem	19/07/2022	19/07/2022
									A070504	Premature Discharge of Battery	19/07/2022	19/07/2022
							F1905	Device Revision or Replacement	A0705	Battery Problem	19/07/2022	19/07/2022
									A070504	Premature Discharge of Battery	19/07/2022	19/07/2022
6641	2022	All other reportable incidents	36241	Pacemaker/defibrillator lead	CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A05	Mechanical Problem	04/05/2022	04/05/2022
									A0709	Device Sensing Problem	04/05/2022	04/05/2022
							F1908	Prolonged Surgery	A05	Mechanical Problem	04/05/2022	04/05/2022
									A0709	Device Sensing Problem	04/05/2022	04/05/2022
6642	2021	All other reportable incidents	47205	Cardiac pulse generator programmer	CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A11	Computer Software Problem		

									A1104	Computer Operating System Problem		
							F26	No Health Consequences or Impact	A11	Computer Software Problem		
									A1104	Computer Operating System Problem		
6643	2021	All other reportable incidents	17882	Physiologic-monitoring defibrillation system	CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A071301	Failure to Deliver Shock/Stimulation		
							F27	Problem Identified During Non-Clinical Procedure	A071301	Failure to Deliver Shock/Stimulation		
6644	2021	All other reportable incidents	35852	Single-chamber implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E2104	Electric Shock	A070909	Oversensing		
									A071302	Inappropriate/Inadequate Shock/Stimulation		
							E2403	No Clinical Signs, Symptoms or Conditions	A070909	Oversensing		
									A071302	Inappropriate/Inadequate Shock/Stimulation		
							F26	No Health Consequences or Impact	A070909	Oversensing		
									A071302	Inappropriate/Inadequate Shock/Stimulation		
6645	2021	All other reportable incidents	35852	Single-chamber implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem	18/10/2021	18/10/2021
									A070504	Premature Discharge of Battery	18/10/2021	18/10/2021
							F08	Hospitalization or Prolonged Hospitalization	A0705	Battery Problem	18/10/2021	18/10/2021
									A070504	Premature Discharge of Battery	18/10/2021	18/10/2021
							F1905	Device Revision or Replacement	A0705	Battery Problem	18/10/2021	18/10/2021
									A070504	Premature Discharge of Battery	18/10/2021	18/10/2021
6646	2021	All other reportable incidents	47265	Dual-chamber implantable pacemaker, rate-responsive	CT141	Electrodes/Leads and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem	23/09/2021	23/09/2021
									A070504	Premature Discharge of Battery	23/09/2021	23/09/2021
							F08	Hospitalization or Prolonged Hospitalization	A0705	Battery Problem	23/09/2021	23/09/2021
									A070504	Premature Discharge of Battery	23/09/2021	23/09/2021
							F1903	Device Explantation	A0705	Battery Problem	23/09/2021	23/09/2021
									A070504	Premature Discharge of Battery	23/09/2021	23/09/2021
					CT145	Cardiovascular devices	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem	23/09/2021	23/09/2021
									A070504	Premature Discharge of Battery	23/09/2021	23/09/2021
							F08	Hospitalization or Prolonged Hospitalization	A0705	Battery Problem	23/09/2021	23/09/2021
									A070504	Premature Discharge of Battery	23/09/2021	23/09/2021
							F1903	Device Explantation	A0705	Battery Problem	23/09/2021	23/09/2021
									A070504	Premature Discharge of Battery	23/09/2021	23/09/2021
					CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem	23/09/2021	23/09/2021
									A070504	Premature Discharge of Battery	23/09/2021	23/09/2021
							F08	Hospitalization or Prolonged Hospitalization	A0705	Battery Problem	23/09/2021	23/09/2021
									A070504	Premature Discharge of Battery	23/09/2021	23/09/2021
							F1903	Device Explantation	A0705	Battery Problem	23/09/2021	23/09/2021
									A070504	Premature Discharge of Battery	23/09/2021	23/09/2021
					CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem	23/09/2021	23/09/2021
									A070504	Premature Discharge of Battery	23/09/2021	23/09/2021
							F08	Hospitalization or Prolonged Hospitalization	A0705	Battery Problem	23/09/2021	23/09/2021
									A070504	Premature Discharge of Battery	23/09/2021	23/09/2021
							F1903	Device Explantation	A0705	Battery Problem	23/09/2021	23/09/2021
									A070504	Premature Discharge of Battery	23/09/2021	23/09/2021
					CT233	Surgical	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem	23/09/2021	23/09/2021
									A070504	Premature Discharge of Battery	23/09/2021	23/09/2021
							F08	Hospitalization or Prolonged Hospitalization	A0705	Battery Problem	23/09/2021	23/09/2021
									A070504	Premature Discharge of Battery	23/09/2021	23/09/2021
							F1903	Device Explantation	A0705	Battery Problem	23/09/2021	23/09/2021
									A070504	Premature Discharge of Battery	23/09/2021	23/09/2021
					CT2406	Implantable	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem	23/09/2021	23/09/2021
									A070504	Premature Discharge of Battery	23/09/2021	23/09/2021
							F08	Hospitalization or Prolonged Hospitalization	A0705	Battery Problem	23/09/2021	23/09/2021
									A070504	Premature Discharge of Battery	23/09/2021	23/09/2021
							F1903	Device Explantation	A0705	Battery Problem	23/09/2021	23/09/2021
									A070504	Premature Discharge of Battery	23/09/2021	23/09/2021
					CT2407	Active implantable	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem	23/09/2021	23/09/2021
									A070504	Premature Discharge of Battery	23/09/2021	23/09/2021
							F08	Hospitalization or Prolonged Hospitalization	A0705	Battery Problem	23/09/2021	23/09/2021
									A070504	Premature Discharge of Battery	23/09/2021	23/09/2021
							F1903	Device Explantation	A0705	Battery Problem	23/09/2021	23/09/2021
									A070504	Premature Discharge of Battery	23/09/2021	23/09/2021
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem	23/09/2021	23/09/2021
									A070504	Premature Discharge of Battery	23/09/2021	23/09/2021
							F08	Hospitalization or Prolonged Hospitalization	A0705	Battery Problem	23/09/2021	23/09/2021
									A070504	Premature Discharge of Battery	23/09/2021	23/09/2021
							F1903	Device Explantation	A0705	Battery Problem	23/09/2021	23/09/2021
									A070504	Premature Discharge of Battery	23/09/2021	23/09/2021
					CT2727	Active implant devices/ancillaries	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem	23/09/2021	23/09/2021
									A070504	Premature Discharge of Battery	23/09/2021	23/09/2021
							F08	Hospitalization or Prolonged Hospitalization	A0705	Battery Problem	23/09/2021	23/09/2021
									A070504	Premature Discharge of Battery	23/09/2021	23/09/2021
							F1903	Device Explantation	A0705	Battery Problem	23/09/2021	23/09/2021
									A070504	Premature Discharge of Battery	23/09/2021	23/09/2021
					CT2750	Implantable cardiac electrical stimulators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem	23/09/2021	23/09/2021
									A070504	Premature Discharge of Battery	23/09/2021	23/09/2021
							F08	Hospitalization or Prolonged Hospitalization	A0705	Battery Problem	23/09/2021	23/09/2021
									A070504	Premature Discharge of Battery	23/09/2021	23/09/2021
							F1903	Device Explantation	A0705	Battery Problem	23/09/2021	23/09/2021
									A070504	Premature Discharge of Battery	23/09/2021	23/09/2021
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem	23/09/2021	23/09/2021
									A070504	Premature Discharge of Battery	23/09/2021	23/09/2021
							F08	Hospitalization or Prolonged Hospitalization	A0705	Battery Problem	23/09/2021	23/09/2021
									A070504	Premature Discharge of Battery	23/09/2021	23/09/2021
							F1903	Device Explantation	A0705	Battery Problem	23/09/2021	23/09/2021
									A070504	Premature Discharge of Battery	23/09/2021	23/09/2021
					CT2985	Active	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem	23/09/2021	23/09/2021
									A070504	Premature Discharge of Battery	23/09/2021	23/09/2021
							F08	Hospitalization or Prolonged Hospitalization	A0705	Battery Problem	23/09/2021	23/09/2021
									A070504	Premature Discharge of Battery	23/09/2021	23/09/2021
							F1903	Device Explantation	A0705	Battery Problem	23/09/2021	23/09/2021

									A070504	Premature Discharge of Battery	23/09/2021	23/09/2021
					CT3001	Implantable metal	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem	23/09/2021	23/09/2021
									A070504	Premature Discharge of Battery	23/09/2021	23/09/2021
							F08	Hospitalization or Prolonged Hospitalization	A0705	Battery Problem	23/09/2021	23/09/2021
									A070504	Premature Discharge of Battery	23/09/2021	23/09/2021
							F1903	Device Explantation	A0705	Battery Problem	23/09/2021	23/09/2021
									A070504	Premature Discharge of Battery	23/09/2021	23/09/2021
					CT3002	Implantable synthetic polymer	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem	23/09/2021	23/09/2021
									A070504	Premature Discharge of Battery	23/09/2021	23/09/2021
							F08	Hospitalization or Prolonged Hospitalization	A0705	Battery Problem	23/09/2021	23/09/2021
									A070504	Premature Discharge of Battery	23/09/2021	23/09/2021
							F1903	Device Explantation	A0705	Battery Problem	23/09/2021	23/09/2021
									A070504	Premature Discharge of Battery	23/09/2021	23/09/2021
					CT3006	Implantable inorganic material	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem	23/09/2021	23/09/2021
									A070504	Premature Discharge of Battery	23/09/2021	23/09/2021
							F08	Hospitalization or Prolonged Hospitalization	A0705	Battery Problem	23/09/2021	23/09/2021
									A070504	Premature Discharge of Battery	23/09/2021	23/09/2021
							F1903	Device Explantation	A0705	Battery Problem	23/09/2021	23/09/2021
									A070504	Premature Discharge of Battery	23/09/2021	23/09/2021
					CT314	Electrically powered	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem	23/09/2021	23/09/2021
									A070504	Premature Discharge of Battery	23/09/2021	23/09/2021
							F08	Hospitalization or Prolonged Hospitalization	A0705	Battery Problem	23/09/2021	23/09/2021
									A070504	Premature Discharge of Battery	23/09/2021	23/09/2021
							F1903	Device Explantation	A0705	Battery Problem	23/09/2021	23/09/2021
									A070504	Premature Discharge of Battery	23/09/2021	23/09/2021
					CT321	Long-term surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem	23/09/2021	23/09/2021
									A070504	Premature Discharge of Battery	23/09/2021	23/09/2021
							F08	Hospitalization or Prolonged Hospitalization	A0705	Battery Problem	23/09/2021	23/09/2021
									A070504	Premature Discharge of Battery	23/09/2021	23/09/2021
							F1903	Device Explantation	A0705	Battery Problem	23/09/2021	23/09/2021
									A070504	Premature Discharge of Battery	23/09/2021	23/09/2021
					CT449	Generators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem	23/09/2021	23/09/2021
									A070504	Premature Discharge of Battery	23/09/2021	23/09/2021
							F08	Hospitalization or Prolonged Hospitalization	A0705	Battery Problem	23/09/2021	23/09/2021
									A070504	Premature Discharge of Battery	23/09/2021	23/09/2021
							F1903	Device Explantation	A0705	Battery Problem	23/09/2021	23/09/2021
									A070504	Premature Discharge of Battery	23/09/2021	23/09/2021
					CT981	Single use	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem	23/09/2021	23/09/2021
									A070504	Premature Discharge of Battery	23/09/2021	23/09/2021
							F08	Hospitalization or Prolonged Hospitalization	A0705	Battery Problem	23/09/2021	23/09/2021
									A070504	Premature Discharge of Battery	23/09/2021	23/09/2021
							F1903	Device Explantation	A0705	Battery Problem	23/09/2021	23/09/2021
									A070504	Premature Discharge of Battery	23/09/2021	23/09/2021
					CT983	Surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem	23/09/2021	23/09/2021
									A070504	Premature Discharge of Battery	23/09/2021	23/09/2021
							F08	Hospitalization or Prolonged Hospitalization	A0705	Battery Problem	23/09/2021	23/09/2021
									A070504	Premature Discharge of Battery	23/09/2021	23/09/2021
							F1903	Device Explantation	A0705	Battery Problem	23/09/2021	23/09/2021
									A070504	Premature Discharge of Battery	23/09/2021	23/09/2021
6647	2022	All other reportable incidents	17882	Physiologic-monitoring defibrillation system	CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A090206	No Display/Image		
							F27	Problem Identified During Non-Clinical Procedure	A090206	No Display/Image		
6648	2022	All other reportable incidents	47270	Cardiac resynchronization therapy implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E0601	Arrhythmia	A070909	Oversensing		
									A070910	Undersensing		
									A0712	Pacing Problem		
									A071202	Inaccurate Synchronization		
							E2403	No Clinical Signs, Symptoms or Conditions	A070909	Oversensing		
									A070910	Undersensing		
									A0712	Pacing Problem		
									A071202	Inaccurate Synchronization		
							F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A070909	Oversensing		
									A070910	Undersensing		
									A0712	Pacing Problem		
									A071202	Inaccurate Synchronization		
6649	2022	All other reportable incidents	35852	Single-chamber implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem	06/12/2022	06/12/2022
									A070504	Premature Discharge of Battery	06/12/2022	06/12/2022
							F08	Hospitalization or Prolonged Hospitalization	A0705	Battery Problem	06/12/2022	06/12/2022
									A070504	Premature Discharge of Battery	06/12/2022	06/12/2022
							F1905	Device Revision or Replacement	A0705	Battery Problem	06/12/2022	06/12/2022
									A070504	Premature Discharge of Battery	06/12/2022	06/12/2022
6650	2022	All other reportable incidents	17882	Physiologic-monitoring defibrillation system	CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A0719	Unexpected Shutdown		
							F27	Problem Identified During Non-Clinical Procedure	A0719	Unexpected Shutdown		
6651	2022	All other reportable incidents	35852	Single-chamber implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem	06/07/2022	06/07/2022
									A070504	Premature Discharge of Battery	06/07/2022	06/07/2022
							F08	Hospitalization or Prolonged Hospitalization	A0705	Battery Problem	06/07/2022	06/07/2022
									A070504	Premature Discharge of Battery	06/07/2022	06/07/2022
							F1905	Device Revision or Replacement	A0705	Battery Problem	06/07/2022	06/07/2022
									A070504	Premature Discharge of Battery	06/07/2022	06/07/2022
6652	2022	All other reportable incidents	37265	Dual-chamber implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E2104	Electric Shock	A070909	Oversensing		
									A0712	Pacing Problem		
									A071302	Inappropriate/Inadequate Shock/Stimulation		
							E2403	No Clinical Signs, Symptoms or Conditions	A070909	Oversensing		
									A0712	Pacing Problem		
									A071302	Inappropriate/Inadequate Shock/Stimulation		
							F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A070909	Oversensing		
									A0712	Pacing Problem		
									A071302	Inappropriate/Inadequate Shock/Stimulation		
6653	2021	All other reportable incidents	47270	Cardiac resynchronization therapy implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E0601	Arrhythmia	A071201	Failure to Convert Rhythm		
							E2403	No Clinical Signs, Symptoms or Conditions	A071201	Failure to Convert Rhythm		

							F26	No Health Consequences or Impact	A071201	Failure to Convert Rhythm		
6654	2021	All other reportable incidents	35223	Endocardial/interventricular septal pacing lead	CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A0401	Break		
									A070909	Oversensing		
							F1901	Additional Surgery	A0401	Break		
									A070909	Oversensing		
6655	2021	All other reportable incidents	47265	Dual-chamber implantable pacemaker, rate-responsive	CT141	Electrodes/Leads and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	30/09/2021	30/09/2021
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	30/09/2021	30/09/2021
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	30/09/2021	30/09/2021
					CT145	Cardiovascular devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	30/09/2021	30/09/2021
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	30/09/2021	30/09/2021
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	30/09/2021	30/09/2021
					CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	30/09/2021	30/09/2021
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	30/09/2021	30/09/2021
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	30/09/2021	30/09/2021
					CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	30/09/2021	30/09/2021
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	30/09/2021	30/09/2021
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	30/09/2021	30/09/2021
					CT233	Surgical	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	30/09/2021	30/09/2021
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	30/09/2021	30/09/2021
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	30/09/2021	30/09/2021
					CT2406	Implantable	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	30/09/2021	30/09/2021
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	30/09/2021	30/09/2021
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	30/09/2021	30/09/2021
					CT2407	Active implantable	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	30/09/2021	30/09/2021
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	30/09/2021	30/09/2021
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	30/09/2021	30/09/2021
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	30/09/2021	30/09/2021
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	30/09/2021	30/09/2021
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	30/09/2021	30/09/2021
					CT2727	Active implant devices/ancillaries	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	30/09/2021	30/09/2021
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	30/09/2021	30/09/2021
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	30/09/2021	30/09/2021
					CT2750	Implantable cardiac electrical stimulators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	30/09/2021	30/09/2021
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	30/09/2021	30/09/2021
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	30/09/2021	30/09/2021
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	30/09/2021	30/09/2021
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	30/09/2021	30/09/2021
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	30/09/2021	30/09/2021
					CT2985	Active	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	30/09/2021	30/09/2021
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	30/09/2021	30/09/2021
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	30/09/2021	30/09/2021
					CT3001	Implantable metal	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	30/09/2021	30/09/2021
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	30/09/2021	30/09/2021
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	30/09/2021	30/09/2021
					CT3002	Implantable synthetic polymer	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	30/09/2021	30/09/2021
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	30/09/2021	30/09/2021
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	30/09/2021	30/09/2021
					CT3006	Implantable inorganic material	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	30/09/2021	30/09/2021
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	30/09/2021	30/09/2021
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	30/09/2021	30/09/2021
					CT314	Electrically powered	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	30/09/2021	30/09/2021
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	30/09/2021	30/09/2021
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	30/09/2021	30/09/2021
					CT321	Long-term surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	30/09/2021	30/09/2021
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	30/09/2021	30/09/2021
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	30/09/2021	30/09/2021
					CT449	Generators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	30/09/2021	30/09/2021
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	30/09/2021	30/09/2021
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	30/09/2021	30/09/2021
					CT981	Single use	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	30/09/2021	30/09/2021
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	30/09/2021	30/09/2021
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	30/09/2021	30/09/2021
					CT983	Surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	30/09/2021	30/09/2021
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	30/09/2021	30/09/2021
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	30/09/2021	30/09/2021
6656	2021	All other reportable incidents	36241	Pacemaker/defibrillator lead	CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A04	Material Integrity Problem	17/09/2021	17/09/2021
									A072201	High Impedance	17/09/2021	17/09/2021
							F08	Hospitalization or Prolonged Hospitalization	A04	Material Integrity Problem	17/09/2021	17/09/2021
									A072201	High Impedance	17/09/2021	17/09/2021
							F1903	Device Explantation	A04	Material Integrity Problem	17/09/2021	17/09/2021
									A072201	High Impedance	17/09/2021	17/09/2021
6657	2021	All other reportable incidents	37265	Dual-chamber implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
6658	2022	All other reportable incidents	35223	Endocardial/interventricular septal pacing lead	CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070901	Decreased Sensitivity		
									A070909	Oversensing		
							F12	Serious Injury/ Illness/ Impairment	A070901	Decreased Sensitivity		
									A070909	Oversensing		
							F23	Unexpected Medical Intervention	A070901	Decreased Sensitivity		
									A070909	Oversensing		
6659	2022	All other reportable incidents	35852	Single-chamber implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem	16/06/2022	16/06/2022
									A070504	Premature Discharge of Battery	16/06/2022	16/06/2022
							F08	Hospitalization or Prolonged Hospitalization	A0705	Battery Problem	16/06/2022	16/06/2022
									A070504	Premature Discharge of Battery	16/06/2022	16/06/2022
							F1905	Device Revision or Replacement	A0705	Battery Problem	16/06/2022	16/06/2022
									A070504	Premature Discharge of Battery	16/06/2022	16/06/2022
6660	2022	All other reportable incidents	47265	Dual-chamber implantable pacemaker, rate-responsive	CT141	Electrodes/Leads and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A0705	Battery Problem		

									A070504	Premature Discharge of Battery			
					CT145	Cardiovascular devices	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem			
									A070504	Premature Discharge of Battery			
							F26	No Health Consequences or Impact	A0705	Battery Problem			
									A070504	Premature Discharge of Battery			
					CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem			
									A070504	Premature Discharge of Battery			
							F26	No Health Consequences or Impact	A0705	Battery Problem			
									A070504	Premature Discharge of Battery			
					CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem			
									A070504	Premature Discharge of Battery			
							F26	No Health Consequences or Impact	A0705	Battery Problem			
									A070504	Premature Discharge of Battery			
					CT233	Surgical	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem			
									A070504	Premature Discharge of Battery			
							F26	No Health Consequences or Impact	A0705	Battery Problem			
									A070504	Premature Discharge of Battery			
					CT2406	Implantable	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem			
									A070504	Premature Discharge of Battery			
							F26	No Health Consequences or Impact	A0705	Battery Problem			
									A070504	Premature Discharge of Battery			
					CT2407	Active implantable	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem			
									A070504	Premature Discharge of Battery			
							F26	No Health Consequences or Impact	A0705	Battery Problem			
									A070504	Premature Discharge of Battery			
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem			
									A070504	Premature Discharge of Battery			
							F26	No Health Consequences or Impact	A0705	Battery Problem			
									A070504	Premature Discharge of Battery			
					CT2727	Active implant devices/ancillaries	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem			
									A070504	Premature Discharge of Battery			
							F26	No Health Consequences or Impact	A0705	Battery Problem			
									A070504	Premature Discharge of Battery			
					CT2750	Implantable cardiac electrical stimulators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem			
									A070504	Premature Discharge of Battery			
							F26	No Health Consequences or Impact	A0705	Battery Problem			
									A070504	Premature Discharge of Battery			
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem			
									A070504	Premature Discharge of Battery			
							F26	No Health Consequences or Impact	A0705	Battery Problem			
									A070504	Premature Discharge of Battery			
					CT2985	Active	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem			
									A070504	Premature Discharge of Battery			
							F26	No Health Consequences or Impact	A0705	Battery Problem			
									A070504	Premature Discharge of Battery			
					CT3001	Implantable metal	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem			
									A070504	Premature Discharge of Battery			
							F26	No Health Consequences or Impact	A0705	Battery Problem			
									A070504	Premature Discharge of Battery			
					CT3002	Implantable synthetic polymer	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem			
									A070504	Premature Discharge of Battery			
							F26	No Health Consequences or Impact	A0705	Battery Problem			
									A070504	Premature Discharge of Battery			
					CT3006	Implantable inorganic material	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem			
									A070504	Premature Discharge of Battery			
							F26	No Health Consequences or Impact	A0705	Battery Problem			
									A070504	Premature Discharge of Battery			
					CT314	Electrically powered	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem			
									A070504	Premature Discharge of Battery			
							F26	No Health Consequences or Impact	A0705	Battery Problem			
									A070504	Premature Discharge of Battery			
					CT321	Long-term surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem			
									A070504	Premature Discharge of Battery			
							F26	No Health Consequences or Impact	A0705	Battery Problem			
									A070504	Premature Discharge of Battery			
					CT449	Generators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem			
									A070504	Premature Discharge of Battery			
							F26	No Health Consequences or Impact	A0705	Battery Problem			
									A070504	Premature Discharge of Battery			
					CT981	Single use	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem			
									A070504	Premature Discharge of Battery			
							F26	No Health Consequences or Impact	A0705	Battery Problem			
									A070504	Premature Discharge of Battery			
					CT983	Surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem			
									A070504	Premature Discharge of Battery			
							F26	No Health Consequences or Impact	A0705	Battery Problem			
									A070504	Premature Discharge of Battery			
6661	2022	All other reportable incidents	17882	Physiologic-monitoring defibrillation system	CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070803	Failure to Power Up			
							F27	Problem Identified During Non-Clinical Procedure	A070803	Failure to Power Up			
6662	2022	All other reportable incidents	47265	Dual-chamber implantable pacemaker, rate-responsive	CT141	Electrodes/Leads and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	12/07/2022	12/07/2022
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	12/07/2022	12/07/2022
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	12/07/2022	12/07/2022
					CT145	Cardiovascular devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	12/07/2022	12/07/2022
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	12/07/2022	12/07/2022
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	12/07/2022	12/07/2022
					CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	12/07/2022	12/07/2022
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	12/07/2022	12/07/2022
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	12/07/2022	12/07/2022

					CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	12/07/2022	12/07/2022
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	12/07/2022	12/07/2022
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	12/07/2022	12/07/2022
					CT233	Surgical	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	12/07/2022	12/07/2022
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	12/07/2022	12/07/2022
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	12/07/2022	12/07/2022
					CT2406	Implantable	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	12/07/2022	12/07/2022
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	12/07/2022	12/07/2022
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	12/07/2022	12/07/2022
					CT2407	Active implantable	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	12/07/2022	12/07/2022
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	12/07/2022	12/07/2022
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	12/07/2022	12/07/2022
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	12/07/2022	12/07/2022
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	12/07/2022	12/07/2022
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	12/07/2022	12/07/2022
					CT2727	Active implant devices/ancillaries	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	12/07/2022	12/07/2022
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	12/07/2022	12/07/2022
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	12/07/2022	12/07/2022
					CT2750	Implantable cardiac electrical stimulators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	12/07/2022	12/07/2022
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	12/07/2022	12/07/2022
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	12/07/2022	12/07/2022
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	12/07/2022	12/07/2022
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	12/07/2022	12/07/2022
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	12/07/2022	12/07/2022
					CT2985	Active	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	12/07/2022	12/07/2022
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	12/07/2022	12/07/2022
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	12/07/2022	12/07/2022
					CT3001	Implantable metal	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	12/07/2022	12/07/2022
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	12/07/2022	12/07/2022
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	12/07/2022	12/07/2022
					CT3002	Implantable synthetic polymer	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	12/07/2022	12/07/2022
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	12/07/2022	12/07/2022
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	12/07/2022	12/07/2022
					CT3006	Implantable inorganic material	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	12/07/2022	12/07/2022
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	12/07/2022	12/07/2022
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	12/07/2022	12/07/2022
					CT314	Electrically powered	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	12/07/2022	12/07/2022
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	12/07/2022	12/07/2022
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	12/07/2022	12/07/2022
					CT321	Long-term surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	12/07/2022	12/07/2022
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	12/07/2022	12/07/2022
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	12/07/2022	12/07/2022
					CT449	Generators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	12/07/2022	12/07/2022
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	12/07/2022	12/07/2022
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	12/07/2022	12/07/2022
					CT981	Single use	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	12/07/2022	12/07/2022
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	12/07/2022	12/07/2022
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	12/07/2022	12/07/2022
					CT983	Surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	12/07/2022	12/07/2022
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	12/07/2022	12/07/2022
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	12/07/2022	12/07/2022
6663	2021	All other reportable incidents	47270	Cardiac resynchronization therapy implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E0601	Arrhythmia	A071201	Failure to Convert Rhythm		
							E2403	No Clinical Signs, Symptoms or Conditions	A071201	Failure to Convert Rhythm		
							F26	No Health Consequences or Impact	A071201	Failure to Convert Rhythm		
6664	2022	All other reportable incidents	47267	Single-chamber implantable pacemaker, rate-responsive	CT141	Electrodes/Leads and associated devices	E0112	Dizziness	A070504	Premature Discharge of Battery	12/05/2022	12/05/2022
									A071102	Failure to Interrogate	12/05/2022	12/05/2022
									A071204	No Pacing	12/05/2022	12/05/2022
									A1107	Data Problem	12/05/2022	12/05/2022
							E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	12/05/2022	12/05/2022
									A071102	Failure to Interrogate	12/05/2022	12/05/2022
									A071204	No Pacing	12/05/2022	12/05/2022
									A1107	Data Problem	12/05/2022	12/05/2022
							F1901	Additional Surgery	A070504	Premature Discharge of Battery	12/05/2022	12/05/2022
									A071102	Failure to Interrogate	12/05/2022	12/05/2022
									A071204	No Pacing	12/05/2022	12/05/2022
									A1107	Data Problem	12/05/2022	12/05/2022
					CT145	Cardiovascular devices	E0112	Dizziness	A070504	Premature Discharge of Battery	12/05/2022	12/05/2022
									A071102	Failure to Interrogate	12/05/2022	12/05/2022
									A071204	No Pacing	12/05/2022	12/05/2022
									A1107	Data Problem	12/05/2022	12/05/2022
							E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	12/05/2022	12/05/2022
									A071102	Failure to Interrogate	12/05/2022	12/05/2022
									A071204	No Pacing	12/05/2022	12/05/2022
									A1107	Data Problem	12/05/2022	12/05/2022
							F1901	Additional Surgery	A070504	Premature Discharge of Battery	12/05/2022	12/05/2022
									A071102	Failure to Interrogate	12/05/2022	12/05/2022
									A071204	No Pacing	12/05/2022	12/05/2022
									A1107	Data Problem	12/05/2022	12/05/2022
					CT1684	Cardiac pacemakers and associated devices	E0112	Dizziness	A070504	Premature Discharge of Battery	12/05/2022	12/05/2022
									A071102	Failure to Interrogate	12/05/2022	12/05/2022
									A071204	No Pacing	12/05/2022	12/05/2022
									A1107	Data Problem	12/05/2022	12/05/2022
							E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	12/05/2022	12/05/2022
									A071102	Failure to Interrogate	12/05/2022	12/05/2022
									A071204	No Pacing	12/05/2022	12/05/2022
									A1107	Data Problem	12/05/2022	12/05/2022
							F1901	Additional Surgery	A070504	Premature Discharge of Battery	12/05/2022	12/05/2022
									A071102	Failure to Interrogate	12/05/2022	12/05/2022

								A071204	No Pacing	12/05/2022	12/05/2022
								A1107	Data Problem	12/05/2022	12/05/2022
					CT1685	Cardiac pacemakers	E0112	Dizziness	A070504	Premature Discharge of Battery	12/05/2022	12/05/2022
									A071102	Failure to Interrogate	12/05/2022	12/05/2022
									A071204	No Pacing	12/05/2022	12/05/2022
									A1107	Data Problem	12/05/2022	12/05/2022
							E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	12/05/2022	12/05/2022
									A071102	Failure to Interrogate	12/05/2022	12/05/2022
									A071204	No Pacing	12/05/2022	12/05/2022
									A1107	Data Problem	12/05/2022	12/05/2022
							F1901	Additional Surgery	A070504	Premature Discharge of Battery	12/05/2022	12/05/2022
									A071102	Failure to Interrogate	12/05/2022	12/05/2022
									A071204	No Pacing	12/05/2022	12/05/2022
									A1107	Data Problem	12/05/2022	12/05/2022
					CT233	Surgical	E0112	Dizziness	A070504	Premature Discharge of Battery	12/05/2022	12/05/2022
									A071102	Failure to Interrogate	12/05/2022	12/05/2022
									A071204	No Pacing	12/05/2022	12/05/2022
									A1107	Data Problem	12/05/2022	12/05/2022
							E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	12/05/2022	12/05/2022
									A071102	Failure to Interrogate	12/05/2022	12/05/2022
									A071204	No Pacing	12/05/2022	12/05/2022
									A1107	Data Problem	12/05/2022	12/05/2022
							F1901	Additional Surgery	A070504	Premature Discharge of Battery	12/05/2022	12/05/2022
									A071102	Failure to Interrogate	12/05/2022	12/05/2022
									A071204	No Pacing	12/05/2022	12/05/2022
									A1107	Data Problem	12/05/2022	12/05/2022
					CT2406	Implantable	E0112	Dizziness	A070504	Premature Discharge of Battery	12/05/2022	12/05/2022
									A071102	Failure to Interrogate	12/05/2022	12/05/2022
									A071204	No Pacing	12/05/2022	12/05/2022
									A1107	Data Problem	12/05/2022	12/05/2022
							E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	12/05/2022	12/05/2022
									A071102	Failure to Interrogate	12/05/2022	12/05/2022
									A071204	No Pacing	12/05/2022	12/05/2022
									A1107	Data Problem	12/05/2022	12/05/2022
							F1901	Additional Surgery	A070504	Premature Discharge of Battery	12/05/2022	12/05/2022
									A071102	Failure to Interrogate	12/05/2022	12/05/2022
									A071204	No Pacing	12/05/2022	12/05/2022
									A1107	Data Problem	12/05/2022	12/05/2022
					CT2407	Active implantable	E0112	Dizziness	A070504	Premature Discharge of Battery	12/05/2022	12/05/2022
									A071102	Failure to Interrogate	12/05/2022	12/05/2022
									A071204	No Pacing	12/05/2022	12/05/2022
									A1107	Data Problem	12/05/2022	12/05/2022
							E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	12/05/2022	12/05/2022
									A071102	Failure to Interrogate	12/05/2022	12/05/2022
									A071204	No Pacing	12/05/2022	12/05/2022
									A1107	Data Problem	12/05/2022	12/05/2022
							F1901	Additional Surgery	A070504	Premature Discharge of Battery	12/05/2022	12/05/2022
									A071102	Failure to Interrogate	12/05/2022	12/05/2022
									A071204	No Pacing	12/05/2022	12/05/2022
									A1107	Data Problem	12/05/2022	12/05/2022
					CT257	Implantable cardiac pacemakers	E0112	Dizziness	A070504	Premature Discharge of Battery	12/05/2022	12/05/2022
									A071102	Failure to Interrogate	12/05/2022	12/05/2022
									A071204	No Pacing	12/05/2022	12/05/2022
									A1107	Data Problem	12/05/2022	12/05/2022
							E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	12/05/2022	12/05/2022
									A071102	Failure to Interrogate	12/05/2022	12/05/2022
									A071204	No Pacing	12/05/2022	12/05/2022
									A1107	Data Problem	12/05/2022	12/05/2022
							F1901	Additional Surgery	A070504	Premature Discharge of Battery	12/05/2022	12/05/2022
									A071102	Failure to Interrogate	12/05/2022	12/05/2022
									A071204	No Pacing	12/05/2022	12/05/2022
									A1107	Data Problem	12/05/2022	12/05/2022
					CT2727	Active implant devices/ancillaries	E0112	Dizziness	A070504	Premature Discharge of Battery	12/05/2022	12/05/2022
									A071102	Failure to Interrogate	12/05/2022	12/05/2022
									A071204	No Pacing	12/05/2022	12/05/2022
									A1107	Data Problem	12/05/2022	12/05/2022
							E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	12/05/2022	12/05/2022
									A071102	Failure to Interrogate	12/05/2022	12/05/2022
									A071204	No Pacing	12/05/2022	12/05/2022
									A1107	Data Problem	12/05/2022	12/05/2022
							F1901	Additional Surgery	A070504	Premature Discharge of Battery	12/05/2022	12/05/2022
									A071102	Failure to Interrogate	12/05/2022	12/05/2022
									A071204	No Pacing	12/05/2022	12/05/2022
									A1107	Data Problem	12/05/2022	12/05/2022
					CT2750	Implantable cardiac electrical stimulators and associated devices	E0112	Dizziness	A070504	Premature Discharge of Battery	12/05/2022	12/05/2022
									A071102	Failure to Interrogate	12/05/2022	12/05/2022
									A071204	No Pacing	12/05/2022	12/05/2022
									A1107	Data Problem	12/05/2022	12/05/2022
							E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	12/05/2022	12/05/2022
									A071102	Failure to Interrogate	12/05/2022	12/05/2022
									A071204	No Pacing	12/05/2022	12/05/2022
									A1107	Data Problem	12/05/2022	12/05/2022
							F1901	Additional Surgery	A070504	Premature Discharge of Battery	12/05/2022	12/05/2022
									A071102	Failure to Interrogate	12/05/2022	12/05/2022
									A071204	No Pacing	12/05/2022	12/05/2022
									A1107	Data Problem	12/05/2022	12/05/2022
					CT2866	Implantable cardiac electrical stimulators	E0112	Dizziness	A070504	Premature Discharge of Battery	12/05/2022	12/05/2022
									A071102	Failure to Interrogate	12/05/2022	12/05/2022

								A071204	No Pacing	12/05/2022	12/05/2022
								A1107	Data Problem	12/05/2022	12/05/2022
						E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	12/05/2022	12/05/2022
								A071102	Failure to Interrogate	12/05/2022	12/05/2022
								A071204	No Pacing	12/05/2022	12/05/2022
								A1107	Data Problem	12/05/2022	12/05/2022
						F1901	Additional Surgery	A070504	Premature Discharge of Battery	12/05/2022	12/05/2022
								A071102	Failure to Interrogate	12/05/2022	12/05/2022
								A071204	No Pacing	12/05/2022	12/05/2022
								A1107	Data Problem	12/05/2022	12/05/2022
					CT2985	Active	E0112	Dizziness	A070504	Premature Discharge of Battery	12/05/2022	12/05/2022
									A071102	Failure to Interrogate	12/05/2022	12/05/2022
									A071204	No Pacing	12/05/2022	12/05/2022
									A1107	Data Problem	12/05/2022	12/05/2022
						E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	12/05/2022	12/05/2022
								A071102	Failure to Interrogate	12/05/2022	12/05/2022
								A071204	No Pacing	12/05/2022	12/05/2022
								A1107	Data Problem	12/05/2022	12/05/2022
						F1901	Additional Surgery	A070504	Premature Discharge of Battery	12/05/2022	12/05/2022
								A071102	Failure to Interrogate	12/05/2022	12/05/2022
								A071204	No Pacing	12/05/2022	12/05/2022
								A1107	Data Problem	12/05/2022	12/05/2022
					CT3001	Implantable metal	E0112	Dizziness	A070504	Premature Discharge of Battery	12/05/2022	12/05/2022
									A071102	Failure to Interrogate	12/05/2022	12/05/2022
									A071204	No Pacing	12/05/2022	12/05/2022
									A1107	Data Problem	12/05/2022	12/05/2022
						E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	12/05/2022	12/05/2022
								A071102	Failure to Interrogate	12/05/2022	12/05/2022
								A071204	No Pacing	12/05/2022	12/05/2022
								A1107	Data Problem	12/05/2022	12/05/2022
						F1901	Additional Surgery	A070504	Premature Discharge of Battery	12/05/2022	12/05/2022
								A071102	Failure to Interrogate	12/05/2022	12/05/2022
								A071204	No Pacing	12/05/2022	12/05/2022
								A1107	Data Problem	12/05/2022	12/05/2022
					CT3002	Implantable synthetic polymer	E0112	Dizziness	A070504	Premature Discharge of Battery	12/05/2022	12/05/2022
									A071102	Failure to Interrogate	12/05/2022	12/05/2022
									A071204	No Pacing	12/05/2022	12/05/2022
									A1107	Data Problem	12/05/2022	12/05/2022
						E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	12/05/2022	12/05/2022
								A071102	Failure to Interrogate	12/05/2022	12/05/2022
								A071204	No Pacing	12/05/2022	12/05/2022
								A1107	Data Problem	12/05/2022	12/05/2022
						F1901	Additional Surgery	A070504	Premature Discharge of Battery	12/05/2022	12/05/2022
								A071102	Failure to Interrogate	12/05/2022	12/05/2022
								A071204	No Pacing	12/05/2022	12/05/2022
								A1107	Data Problem	12/05/2022	12/05/2022
					CT3006	Implantable inorganic material	E0112	Dizziness	A070504	Premature Discharge of Battery	12/05/2022	12/05/2022
									A071102	Failure to Interrogate	12/05/2022	12/05/2022
									A071204	No Pacing	12/05/2022	12/05/2022
									A1107	Data Problem	12/05/2022	12/05/2022
						E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	12/05/2022	12/05/2022
								A071102	Failure to Interrogate	12/05/2022	12/05/2022
								A071204	No Pacing	12/05/2022	12/05/2022
								A1107	Data Problem	12/05/2022	12/05/2022
						F1901	Additional Surgery	A070504	Premature Discharge of Battery	12/05/2022	12/05/2022
								A071102	Failure to Interrogate	12/05/2022	12/05/2022
								A071204	No Pacing	12/05/2022	12/05/2022
								A1107	Data Problem	12/05/2022	12/05/2022
					CT314	Electrically powered	E0112	Dizziness	A070504	Premature Discharge of Battery	12/05/2022	12/05/2022
									A071102	Failure to Interrogate	12/05/2022	12/05/2022
									A071204	No Pacing	12/05/2022	12/05/2022
									A1107	Data Problem	12/05/2022	12/05/2022
						E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	12/05/2022	12/05/2022
								A071102	Failure to Interrogate	12/05/2022	12/05/2022
								A071204	No Pacing	12/05/2022	12/05/2022
								A1107	Data Problem	12/05/2022	12/05/2022
						F1901	Additional Surgery	A070504	Premature Discharge of Battery	12/05/2022	12/05/2022
								A071102	Failure to Interrogate	12/05/2022	12/05/2022
								A071204	No Pacing	12/05/2022	12/05/2022
								A1107	Data Problem	12/05/2022	12/05/2022
					CT321	Long-term surgical invasive	E0112	Dizziness	A070504	Premature Discharge of Battery	12/05/2022	12/05/2022
									A071102	Failure to Interrogate	12/05/2022	12/05/2022
									A071204	No Pacing	12/05/2022	12/05/2022
									A1107	Data Problem	12/05/2022	12/05/2022
						E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	12/05/2022	12/05/2022
								A071102	Failure to Interrogate	12/05/2022	12/05/2022
								A071204	No Pacing	12/05/2022	12/05/2022
								A1107	Data Problem	12/05/2022	12/05/2022
						F1901	Additional Surgery	A070504	Premature Discharge of Battery	12/05/2022	12/05/2022
								A071102	Failure to Interrogate	12/05/2022	12/05/2022
								A071204	No Pacing	12/05/2022	12/05/2022
								A1107	Data Problem	12/05/2022	12/05/2022
					CT449	Generators and associated devices	E0112	Dizziness	A070504	Premature Discharge of Battery	12/05/2022	12/05/2022
									A071102	Failure to Interrogate	12/05/2022	12/05/2022
									A071204	No Pacing	12/05/2022	12/05/2022
									A1107	Data Problem	12/05/2022	12/05/2022
						E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	12/05/2022	12/05/2022
								A071102	Failure to Interrogate	12/05/2022	12/05/2022

								A071204	No Pacing	12/05/2022	12/05/2022
								A1107	Data Problem	12/05/2022	12/05/2022
						F1901	Additional Surgery	A070504	Premature Discharge of Battery	12/05/2022	12/05/2022
								A071102	Failure to Interrogate	12/05/2022	12/05/2022
								A071204	No Pacing	12/05/2022	12/05/2022
								A1107	Data Problem	12/05/2022	12/05/2022
				CT981	Single use	E0112	Dizziness	A070504	Premature Discharge of Battery	12/05/2022	12/05/2022
								A071102	Failure to Interrogate	12/05/2022	12/05/2022
								A071204	No Pacing	12/05/2022	12/05/2022
								A1107	Data Problem	12/05/2022	12/05/2022
						E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	12/05/2022	12/05/2022
								A071102	Failure to Interrogate	12/05/2022	12/05/2022
								A071204	No Pacing	12/05/2022	12/05/2022
								A1107	Data Problem	12/05/2022	12/05/2022
						F1901	Additional Surgery	A070504	Premature Discharge of Battery	12/05/2022	12/05/2022
								A071102	Failure to Interrogate	12/05/2022	12/05/2022
								A071204	No Pacing	12/05/2022	12/05/2022
								A1107	Data Problem	12/05/2022	12/05/2022
				CT983	Surgical invasive	E0112	Dizziness	A070504	Premature Discharge of Battery	12/05/2022	12/05/2022
								A071102	Failure to Interrogate	12/05/2022	12/05/2022
								A071204	No Pacing	12/05/2022	12/05/2022
								A1107	Data Problem	12/05/2022	12/05/2022
						E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	12/05/2022	12/05/2022
								A071102	Failure to Interrogate	12/05/2022	12/05/2022
								A071204	No Pacing	12/05/2022	12/05/2022
								A1107	Data Problem	12/05/2022	12/05/2022
						F1901	Additional Surgery	A070504	Premature Discharge of Battery	12/05/2022	12/05/2022
								A071102	Failure to Interrogate	12/05/2022	12/05/2022
								A071204	No Pacing	12/05/2022	12/05/2022
								A1107	Data Problem	12/05/2022	12/05/2022
6665	2022	All other reportable incidents	47265	Dual-chamber implantable pacemaker, rate-responsive	CT141	Electrodes/Leads and associated devices	E0601	Arrhythmia	A071205	Pacemaker Found in Back-Up Mode	10/05/2022	10/05/2022
									A2303	Improper or Incorrect Procedure or Method	10/05/2022	10/05/2022
							E2403	No Clinical Signs, Symptoms or Conditions	A071205	Pacemaker Found in Back-Up Mode	10/05/2022	10/05/2022
									A2303	Improper or Incorrect Procedure or Method	10/05/2022	10/05/2022
							F1908	Prolonged Surgery	A071205	Pacemaker Found in Back-Up Mode	10/05/2022	10/05/2022
									A2303	Improper or Incorrect Procedure or Method	10/05/2022	10/05/2022
					CT145	Cardiovascular devices	E0601	Arrhythmia	A071205	Pacemaker Found in Back-Up Mode	10/05/2022	10/05/2022
									A2303	Improper or Incorrect Procedure or Method	10/05/2022	10/05/2022
							E2403	No Clinical Signs, Symptoms or Conditions	A071205	Pacemaker Found in Back-Up Mode	10/05/2022	10/05/2022
									A2303	Improper or Incorrect Procedure or Method	10/05/2022	10/05/2022
							F1908	Prolonged Surgery	A071205	Pacemaker Found in Back-Up Mode	10/05/2022	10/05/2022
									A2303	Improper or Incorrect Procedure or Method	10/05/2022	10/05/2022
					CT1684	Cardiac pacemakers and associated devices	E0601	Arrhythmia	A071205	Pacemaker Found in Back-Up Mode	10/05/2022	10/05/2022
									A2303	Improper or Incorrect Procedure or Method	10/05/2022	10/05/2022
							E2403	No Clinical Signs, Symptoms or Conditions	A071205	Pacemaker Found in Back-Up Mode	10/05/2022	10/05/2022
									A2303	Improper or Incorrect Procedure or Method	10/05/2022	10/05/2022
							F1908	Prolonged Surgery	A071205	Pacemaker Found in Back-Up Mode	10/05/2022	10/05/2022
									A2303	Improper or Incorrect Procedure or Method	10/05/2022	10/05/2022
					CT1685	Cardiac pacemakers	E0601	Arrhythmia	A071205	Pacemaker Found in Back-Up Mode	10/05/2022	10/05/2022
									A2303	Improper or Incorrect Procedure or Method	10/05/2022	10/05/2022
							E2403	No Clinical Signs, Symptoms or Conditions	A071205	Pacemaker Found in Back-Up Mode	10/05/2022	10/05/2022
									A2303	Improper or Incorrect Procedure or Method	10/05/2022	10/05/2022
							F1908	Prolonged Surgery	A071205	Pacemaker Found in Back-Up Mode	10/05/2022	10/05/2022
									A2303	Improper or Incorrect Procedure or Method	10/05/2022	10/05/2022
					CT233	Surgical	E0601	Arrhythmia	A071205	Pacemaker Found in Back-Up Mode	10/05/2022	10/05/2022
									A2303	Improper or Incorrect Procedure or Method	10/05/2022	10/05/2022
							E2403	No Clinical Signs, Symptoms or Conditions	A071205	Pacemaker Found in Back-Up Mode	10/05/2022	10/05/2022
									A2303	Improper or Incorrect Procedure or Method	10/05/2022	10/05/2022
							F1908	Prolonged Surgery	A071205	Pacemaker Found in Back-Up Mode	10/05/2022	10/05/2022
									A2303	Improper or Incorrect Procedure or Method	10/05/2022	10/05/2022
					CT2406	Implantable	E0601	Arrhythmia	A071205	Pacemaker Found in Back-Up Mode	10/05/2022	10/05/2022
									A2303	Improper or Incorrect Procedure or Method	10/05/2022	10/05/2022
							E2403	No Clinical Signs, Symptoms or Conditions	A071205	Pacemaker Found in Back-Up Mode	10/05/2022	10/05/2022
									A2303	Improper or Incorrect Procedure or Method	10/05/2022	10/05/2022
							F1908	Prolonged Surgery	A071205	Pacemaker Found in Back-Up Mode	10/05/2022	10/05/2022
									A2303	Improper or Incorrect Procedure or Method	10/05/2022	10/05/2022
					CT2407	Active implantable	E0601	Arrhythmia	A071205	Pacemaker Found in Back-Up Mode	10/05/2022	10/05/2022
									A2303	Improper or Incorrect Procedure or Method	10/05/2022	10/05/2022
							E2403	No Clinical Signs, Symptoms or Conditions	A071205	Pacemaker Found in Back-Up Mode	10/05/2022	10/05/2022
									A2303	Improper or Incorrect Procedure or Method	10/05/2022	10/05/2022
							F1908	Prolonged Surgery	A071205	Pacemaker Found in Back-Up Mode	10/05/2022	10/05/2022
									A2303	Improper or Incorrect Procedure or Method	10/05/2022	10/05/2022
					CT257	Implantable cardiac pacemakers	E0601	Arrhythmia	A071205	Pacemaker Found in Back-Up Mode	10/05/2022	10/05/2022
									A2303	Improper or Incorrect Procedure or Method	10/05/2022	10/05/2022
							E2403	No Clinical Signs, Symptoms or Conditions	A071205	Pacemaker Found in Back-Up Mode	10/05/2022	10/05/2022
									A2303	Improper or Incorrect Procedure or Method	10/05/2022	10/05/2022
							F1908	Prolonged Surgery	A071205	Pacemaker Found in Back-Up Mode	10/05/2022	10/05/2022
									A2303	Improper or Incorrect Procedure or Method	10/05/2022	10/05/2022
					CT2727	Active implant devices/ancillaries	E0601	Arrhythmia	A071205	Pacemaker Found in Back-Up Mode	10/05/2022	10/05/2022
									A2303	Improper or Incorrect Procedure or Method	10/05/2022	10/05/2022
							E2403	No Clinical Signs, Symptoms or Conditions	A071205	Pacemaker Found in Back-Up Mode	10/05/2022	10/05/2022
									A2303	Improper or Incorrect Procedure or Method	10/05/2022	10/05/2022
							F1908	Prolonged Surgery	A071205	Pacemaker Found in Back-Up Mode	10/05/2022	10/05/2022
									A2303	Improper or Incorrect Procedure or Method	10/05/2022	10/05/2022
					CT2750	Implantable cardiac electrical stimulators and associated devices	E0601	Arrhythmia	A071205	Pacemaker Found in Back-Up Mode	10/05/2022	10/05/2022
									A2303	Improper or Incorrect Procedure or Method	10/05/2022	10/05/2022
							E2403	No Clinical Signs, Symptoms or Conditions	A071205	Pacemaker Found in Back-Up Mode	10/05/2022	10/05/2022
									A2303	Improper or Incorrect Procedure or Method	10/05/2022	10/05/2022

							F1908	Prolonged Surgery	A071205	Pacemaker Found in Back-Up Mode	10/05/2022	10/05/2022
									A2303	Improper or Incorrect Procedure or Method	10/05/2022	10/05/2022
					CT2866	Implantable cardiac electrical stimulators	E0601	Arrhythmia	A071205	Pacemaker Found in Back-Up Mode	10/05/2022	10/05/2022
									A2303	Improper or Incorrect Procedure or Method	10/05/2022	10/05/2022
							E2403	No Clinical Signs, Symptoms or Conditions	A071205	Pacemaker Found in Back-Up Mode	10/05/2022	10/05/2022
									A2303	Improper or Incorrect Procedure or Method	10/05/2022	10/05/2022
							F1908	Prolonged Surgery	A071205	Pacemaker Found in Back-Up Mode	10/05/2022	10/05/2022
									A2303	Improper or Incorrect Procedure or Method	10/05/2022	10/05/2022
					CT2985	Active	E0601	Arrhythmia	A071205	Pacemaker Found in Back-Up Mode	10/05/2022	10/05/2022
									A2303	Improper or Incorrect Procedure or Method	10/05/2022	10/05/2022
							E2403	No Clinical Signs, Symptoms or Conditions	A071205	Pacemaker Found in Back-Up Mode	10/05/2022	10/05/2022
									A2303	Improper or Incorrect Procedure or Method	10/05/2022	10/05/2022
							F1908	Prolonged Surgery	A071205	Pacemaker Found in Back-Up Mode	10/05/2022	10/05/2022
									A2303	Improper or Incorrect Procedure or Method	10/05/2022	10/05/2022
					CT3001	Implantable metal	E0601	Arrhythmia	A071205	Pacemaker Found in Back-Up Mode	10/05/2022	10/05/2022
									A2303	Improper or Incorrect Procedure or Method	10/05/2022	10/05/2022
							E2403	No Clinical Signs, Symptoms or Conditions	A071205	Pacemaker Found in Back-Up Mode	10/05/2022	10/05/2022
									A2303	Improper or Incorrect Procedure or Method	10/05/2022	10/05/2022
							F1908	Prolonged Surgery	A071205	Pacemaker Found in Back-Up Mode	10/05/2022	10/05/2022
									A2303	Improper or Incorrect Procedure or Method	10/05/2022	10/05/2022
					CT3002	Implantable synthetic polymer	E0601	Arrhythmia	A071205	Pacemaker Found in Back-Up Mode	10/05/2022	10/05/2022
									A2303	Improper or Incorrect Procedure or Method	10/05/2022	10/05/2022
							E2403	No Clinical Signs, Symptoms or Conditions	A071205	Pacemaker Found in Back-Up Mode	10/05/2022	10/05/2022
									A2303	Improper or Incorrect Procedure or Method	10/05/2022	10/05/2022
							F1908	Prolonged Surgery	A071205	Pacemaker Found in Back-Up Mode	10/05/2022	10/05/2022
									A2303	Improper or Incorrect Procedure or Method	10/05/2022	10/05/2022
					CT3006	Implantable inorganic material	E0601	Arrhythmia	A071205	Pacemaker Found in Back-Up Mode	10/05/2022	10/05/2022
									A2303	Improper or Incorrect Procedure or Method	10/05/2022	10/05/2022
							E2403	No Clinical Signs, Symptoms or Conditions	A071205	Pacemaker Found in Back-Up Mode	10/05/2022	10/05/2022
									A2303	Improper or Incorrect Procedure or Method	10/05/2022	10/05/2022
							F1908	Prolonged Surgery	A071205	Pacemaker Found in Back-Up Mode	10/05/2022	10/05/2022
									A2303	Improper or Incorrect Procedure or Method	10/05/2022	10/05/2022
					CT314	Electrically powered	E0601	Arrhythmia	A071205	Pacemaker Found in Back-Up Mode	10/05/2022	10/05/2022
									A2303	Improper or Incorrect Procedure or Method	10/05/2022	10/05/2022
							E2403	No Clinical Signs, Symptoms or Conditions	A071205	Pacemaker Found in Back-Up Mode	10/05/2022	10/05/2022
									A2303	Improper or Incorrect Procedure or Method	10/05/2022	10/05/2022
							F1908	Prolonged Surgery	A071205	Pacemaker Found in Back-Up Mode	10/05/2022	10/05/2022
									A2303	Improper or Incorrect Procedure or Method	10/05/2022	10/05/2022
					CT321	Long-term surgical invasive	E0601	Arrhythmia	A071205	Pacemaker Found in Back-Up Mode	10/05/2022	10/05/2022
									A2303	Improper or Incorrect Procedure or Method	10/05/2022	10/05/2022
							E2403	No Clinical Signs, Symptoms or Conditions	A071205	Pacemaker Found in Back-Up Mode	10/05/2022	10/05/2022
									A2303	Improper or Incorrect Procedure or Method	10/05/2022	10/05/2022
							F1908	Prolonged Surgery	A071205	Pacemaker Found in Back-Up Mode	10/05/2022	10/05/2022
									A2303	Improper or Incorrect Procedure or Method	10/05/2022	10/05/2022
					CT449	Generators and associated devices	E0601	Arrhythmia	A071205	Pacemaker Found in Back-Up Mode	10/05/2022	10/05/2022
									A2303	Improper or Incorrect Procedure or Method	10/05/2022	10/05/2022
							E2403	No Clinical Signs, Symptoms or Conditions	A071205	Pacemaker Found in Back-Up Mode	10/05/2022	10/05/2022
									A2303	Improper or Incorrect Procedure or Method	10/05/2022	10/05/2022
							F1908	Prolonged Surgery	A071205	Pacemaker Found in Back-Up Mode	10/05/2022	10/05/2022
									A2303	Improper or Incorrect Procedure or Method	10/05/2022	10/05/2022
					CT981	Single use	E0601	Arrhythmia	A071205	Pacemaker Found in Back-Up Mode	10/05/2022	10/05/2022
									A2303	Improper or Incorrect Procedure or Method	10/05/2022	10/05/2022
							E2403	No Clinical Signs, Symptoms or Conditions	A071205	Pacemaker Found in Back-Up Mode	10/05/2022	10/05/2022
									A2303	Improper or Incorrect Procedure or Method	10/05/2022	10/05/2022
							F1908	Prolonged Surgery	A071205	Pacemaker Found in Back-Up Mode	10/05/2022	10/05/2022
									A2303	Improper or Incorrect Procedure or Method	10/05/2022	10/05/2022
					CT983	Surgical invasive	E0601	Arrhythmia	A071205	Pacemaker Found in Back-Up Mode	10/05/2022	10/05/2022
									A2303	Improper or Incorrect Procedure or Method	10/05/2022	10/05/2022
							E2403	No Clinical Signs, Symptoms or Conditions	A071205	Pacemaker Found in Back-Up Mode	10/05/2022	10/05/2022
									A2303	Improper or Incorrect Procedure or Method	10/05/2022	10/05/2022
							F1908	Prolonged Surgery	A071205	Pacemaker Found in Back-Up Mode	10/05/2022	10/05/2022
									A2303	Improper or Incorrect Procedure or Method	10/05/2022	10/05/2022
6666	2022	All other reportable incidents	35852	Single-chamber implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem	06/12/2022	06/12/2022
									A070504	Premature Discharge of Battery	06/12/2022	06/12/2022
							F08	Hospitalization or Prolonged Hospitalization	A0705	Battery Problem	06/12/2022	06/12/2022
									A070504	Premature Discharge of Battery	06/12/2022	06/12/2022
							F1903	Device Explantation	A0705	Battery Problem	06/12/2022	06/12/2022
									A070504	Premature Discharge of Battery	06/12/2022	06/12/2022
6667	2021	All other reportable incidents	36241	Pacemaker/defibrillator lead	CT1684	Cardiac pacemakers and associated devices	E2104	Electric Shock	A070909	Oversensing		
									A071302	Inappropriate/Inadequate Shock/Stimulation		
									A090801	Signal Artifact/Noise		
									A1301	Failure to Read Input Signal		
							E2403	No Clinical Signs, Symptoms or Conditions	A070909	Oversensing		
									A071302	Inappropriate/Inadequate Shock/Stimulation		
									A090801	Signal Artifact/Noise		
									A1301	Failure to Read Input Signal		
							F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A070909	Oversensing		
									A071302	Inappropriate/Inadequate Shock/Stimulation		
									A090801	Signal Artifact/Noise		
									A1301	Failure to Read Input Signal		
6668	2021	All other reportable incidents	17882	Physiologic-monitoring defibrillation system	CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070803	Failure to Power Up		
							F27	Problem Identified During Non-Clinical Procedure	A070803	Failure to Power Up		
6669	2022	All other reportable incidents	47267	Single-chamber implantable pacemaker, rate-responsive	CT141	Electrodes/Leads and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A04	Material Integrity Problem	15/05/2022	15/05/2022
									A070101	Failure to Capture	15/05/2022	15/05/2022
									A070102	High Capture Threshold	15/05/2022	15/05/2022
									A072201	High Impedance	15/05/2022	15/05/2022
							F08	Hospitalization or Prolonged Hospitalization	A04	Material Integrity Problem	15/05/2022	15/05/2022
									A070101	Failure to Capture	15/05/2022	15/05/2022

									A070102	High Capture Threshold	15/05/2022	15/05/2022
									A072201	High Impedance	15/05/2022	15/05/2022
								F1905	Device Revision or Replacement	A04	Material Integrity Problem	15/05/2022	15/05/2022
											A070101	Failure to Capture	15/05/2022	15/05/2022
											A070102	High Capture Threshold	15/05/2022	15/05/2022
											A072201	High Impedance	15/05/2022	15/05/2022
							CT145	Cardiovascular devices	E2403	No Clinical Signs, Symptoms or Conditions	A04	Material Integrity Problem	15/05/2022	15/05/2022
											A070101	Failure to Capture	15/05/2022	15/05/2022
											A070102	High Capture Threshold	15/05/2022	15/05/2022
											A072201	High Impedance	15/05/2022	15/05/2022
									F08	Hospitalization or Prolonged Hospitalization	A04	Material Integrity Problem	15/05/2022	15/05/2022
											A070101	Failure to Capture	15/05/2022	15/05/2022
											A070102	High Capture Threshold	15/05/2022	15/05/2022
											A072201	High Impedance	15/05/2022	15/05/2022
									F1905	Device Revision or Replacement	A04	Material Integrity Problem	15/05/2022	15/05/2022
											A070101	Failure to Capture	15/05/2022	15/05/2022
											A070102	High Capture Threshold	15/05/2022	15/05/2022
											A072201	High Impedance	15/05/2022	15/05/2022
							CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A04	Material Integrity Problem	15/05/2022	15/05/2022
											A070101	Failure to Capture	15/05/2022	15/05/2022
											A070102	High Capture Threshold	15/05/2022	15/05/2022
											A072201	High Impedance	15/05/2022	15/05/2022
									F08	Hospitalization or Prolonged Hospitalization	A04	Material Integrity Problem	15/05/2022	15/05/2022
											A070101	Failure to Capture	15/05/2022	15/05/2022
											A070102	High Capture Threshold	15/05/2022	15/05/2022
											A072201	High Impedance	15/05/2022	15/05/2022
									F1905	Device Revision or Replacement	A04	Material Integrity Problem	15/05/2022	15/05/2022
											A070101	Failure to Capture	15/05/2022	15/05/2022
											A070102	High Capture Threshold	15/05/2022	15/05/2022
											A072201	High Impedance	15/05/2022	15/05/2022
							CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A04	Material Integrity Problem	15/05/2022	15/05/2022
											A070101	Failure to Capture	15/05/2022	15/05/2022
											A070102	High Capture Threshold	15/05/2022	15/05/2022
											A072201	High Impedance	15/05/2022	15/05/2022
									F08	Hospitalization or Prolonged Hospitalization	A04	Material Integrity Problem	15/05/2022	15/05/2022
											A070101	Failure to Capture	15/05/2022	15/05/2022
											A070102	High Capture Threshold	15/05/2022	15/05/2022
											A072201	High Impedance	15/05/2022	15/05/2022
									F1905	Device Revision or Replacement	A04	Material Integrity Problem	15/05/2022	15/05/2022
											A070101	Failure to Capture	15/05/2022	15/05/2022
											A070102	High Capture Threshold	15/05/2022	15/05/2022
											A072201	High Impedance	15/05/2022	15/05/2022
							CT233	Surgical	E2403	No Clinical Signs, Symptoms or Conditions	A04	Material Integrity Problem	15/05/2022	15/05/2022
											A070101	Failure to Capture	15/05/2022	15/05/2022
											A070102	High Capture Threshold	15/05/2022	15/05/2022
											A072201	High Impedance	15/05/2022	15/05/2022
									F08	Hospitalization or Prolonged Hospitalization	A04	Material Integrity Problem	15/05/2022	15/05/2022
											A070101	Failure to Capture	15/05/2022	15/05/2022
											A070102	High Capture Threshold	15/05/2022	15/05/2022
											A072201	High Impedance	15/05/2022	15/05/2022
									F1905	Device Revision or Replacement	A04	Material Integrity Problem	15/05/2022	15/05/2022
											A070101	Failure to Capture	15/05/2022	15/05/2022
											A070102	High Capture Threshold	15/05/2022	15/05/2022
											A072201	High Impedance	15/05/2022	15/05/2022
							CT2406	Implantable	E2403	No Clinical Signs, Symptoms or Conditions	A04	Material Integrity Problem	15/05/2022	15/05/2022
											A070101	Failure to Capture	15/05/2022	15/05/2022
											A070102	High Capture Threshold	15/05/2022	15/05/2022
											A072201	High Impedance	15/05/2022	15/05/2022
									F08	Hospitalization or Prolonged Hospitalization	A04	Material Integrity Problem	15/05/2022	15/05/2022
											A070101	Failure to Capture	15/05/2022	15/05/2022
											A070102	High Capture Threshold	15/05/2022	15/05/2022
											A072201	High Impedance	15/05/2022	15/05/2022
									F1905	Device Revision or Replacement	A04	Material Integrity Problem	15/05/2022	15/05/2022
											A070101	Failure to Capture	15/05/2022	15/05/2022
											A070102	High Capture Threshold	15/05/2022	15/05/2022
											A072201	High Impedance	15/05/2022	15/05/2022
							CT2407	Active implantable	E2403	No Clinical Signs, Symptoms or Conditions	A04	Material Integrity Problem	15/05/2022	15/05/2022
											A070101	Failure to Capture	15/05/2022	15/05/2022
											A070102	High Capture Threshold	15/05/2022	15/05/2022
											A072201	High Impedance	15/05/2022	15/05/2022
									F08	Hospitalization or Prolonged Hospitalization	A04	Material Integrity Problem	15/05/2022	15/05/2022
											A070101	Failure to Capture	15/05/2022	15/05/2022
											A070102	High Capture Threshold	15/05/2022	15/05/2022
											A072201	High Impedance	15/05/2022	15/05/2022
									F1905	Device Revision or Replacement	A04	Material Integrity Problem	15/05/2022	15/05/2022
											A070101	Failure to Capture	15/05/2022	15/05/2022
											A070102	High Capture Threshold	15/05/2022	15/05/2022
											A072201	High Impedance	15/05/2022	15/05/2022
							CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A04	Material Integrity Problem	15/05/2022	15/05/2022
											A070101	Failure to Capture	15/05/2022	15/05/2022
											A070102	High Capture Threshold	15/05/2022	15/05/2022
											A072201	High Impedance	15/05/2022	15/05/2022
									F08	Hospitalization or Prolonged Hospitalization	A04	Material Integrity Problem	15/05/2022	15/05/2022
											A070101	Failure to Capture	15/05/2022	15/05/2022
											A070102	High Capture Threshold	15/05/2022	15/05/2022
											A072201	High Impedance	15/05/2022	15/05/2022
									F1905	Device Revision or Replacement	A04	Material Integrity Problem	15/05/2022	15/05/2022
											A070101	Failure to Capture	15/05/2022	15/05/2022

								A070102	High Capture Threshold	15/05/2022	15/05/2022
								A072201	High Impedance	15/05/2022	15/05/2022
					CT2727	Active implant devices/ancillaries	E2403	No Clinical Signs, Symptoms or Conditions	A04	Material Integrity Problem	15/05/2022	15/05/2022
									A070101	Failure to Capture	15/05/2022	15/05/2022
									A070102	High Capture Threshold	15/05/2022	15/05/2022
									A072201	High Impedance	15/05/2022	15/05/2022
							F08	Hospitalization or Prolonged Hospitalization	A04	Material Integrity Problem	15/05/2022	15/05/2022
									A070101	Failure to Capture	15/05/2022	15/05/2022
									A070102	High Capture Threshold	15/05/2022	15/05/2022
									A072201	High Impedance	15/05/2022	15/05/2022
							F1905	Device Revision or Replacement	A04	Material Integrity Problem	15/05/2022	15/05/2022
									A070101	Failure to Capture	15/05/2022	15/05/2022
									A070102	High Capture Threshold	15/05/2022	15/05/2022
									A072201	High Impedance	15/05/2022	15/05/2022
					CT2750	Implantable cardiac electrical stimulators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A04	Material Integrity Problem	15/05/2022	15/05/2022
									A070101	Failure to Capture	15/05/2022	15/05/2022
									A070102	High Capture Threshold	15/05/2022	15/05/2022
									A072201	High Impedance	15/05/2022	15/05/2022
							F08	Hospitalization or Prolonged Hospitalization	A04	Material Integrity Problem	15/05/2022	15/05/2022
									A070101	Failure to Capture	15/05/2022	15/05/2022
									A070102	High Capture Threshold	15/05/2022	15/05/2022
									A072201	High Impedance	15/05/2022	15/05/2022
							F1905	Device Revision or Replacement	A04	Material Integrity Problem	15/05/2022	15/05/2022
									A070101	Failure to Capture	15/05/2022	15/05/2022
									A070102	High Capture Threshold	15/05/2022	15/05/2022
									A072201	High Impedance	15/05/2022	15/05/2022
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A04	Material Integrity Problem	15/05/2022	15/05/2022
									A070101	Failure to Capture	15/05/2022	15/05/2022
									A070102	High Capture Threshold	15/05/2022	15/05/2022
									A072201	High Impedance	15/05/2022	15/05/2022
							F08	Hospitalization or Prolonged Hospitalization	A04	Material Integrity Problem	15/05/2022	15/05/2022
									A070101	Failure to Capture	15/05/2022	15/05/2022
									A070102	High Capture Threshold	15/05/2022	15/05/2022
									A072201	High Impedance	15/05/2022	15/05/2022
							F1905	Device Revision or Replacement	A04	Material Integrity Problem	15/05/2022	15/05/2022
									A070101	Failure to Capture	15/05/2022	15/05/2022
									A070102	High Capture Threshold	15/05/2022	15/05/2022
									A072201	High Impedance	15/05/2022	15/05/2022
					CT2985	Active	E2403	No Clinical Signs, Symptoms or Conditions	A04	Material Integrity Problem	15/05/2022	15/05/2022
									A070101	Failure to Capture	15/05/2022	15/05/2022
									A070102	High Capture Threshold	15/05/2022	15/05/2022
									A072201	High Impedance	15/05/2022	15/05/2022
							F08	Hospitalization or Prolonged Hospitalization	A04	Material Integrity Problem	15/05/2022	15/05/2022
									A070101	Failure to Capture	15/05/2022	15/05/2022
									A070102	High Capture Threshold	15/05/2022	15/05/2022
									A072201	High Impedance	15/05/2022	15/05/2022
							F1905	Device Revision or Replacement	A04	Material Integrity Problem	15/05/2022	15/05/2022
									A070101	Failure to Capture	15/05/2022	15/05/2022
									A070102	High Capture Threshold	15/05/2022	15/05/2022
									A072201	High Impedance	15/05/2022	15/05/2022
					CT3001	Implantable metal	E2403	No Clinical Signs, Symptoms or Conditions	A04	Material Integrity Problem	15/05/2022	15/05/2022
									A070101	Failure to Capture	15/05/2022	15/05/2022
									A070102	High Capture Threshold	15/05/2022	15/05/2022
									A072201	High Impedance	15/05/2022	15/05/2022
							F08	Hospitalization or Prolonged Hospitalization	A04	Material Integrity Problem	15/05/2022	15/05/2022
									A070101	Failure to Capture	15/05/2022	15/05/2022
									A070102	High Capture Threshold	15/05/2022	15/05/2022
									A072201	High Impedance	15/05/2022	15/05/2022
							F1905	Device Revision or Replacement	A04	Material Integrity Problem	15/05/2022	15/05/2022
									A070101	Failure to Capture	15/05/2022	15/05/2022
									A070102	High Capture Threshold	15/05/2022	15/05/2022
									A072201	High Impedance	15/05/2022	15/05/2022
					CT3002	Implantable synthetic polymer	E2403	No Clinical Signs, Symptoms or Conditions	A04	Material Integrity Problem	15/05/2022	15/05/2022
									A070101	Failure to Capture	15/05/2022	15/05/2022
									A070102	High Capture Threshold	15/05/2022	15/05/2022
									A072201	High Impedance	15/05/2022	15/05/2022
							F08	Hospitalization or Prolonged Hospitalization	A04	Material Integrity Problem	15/05/2022	15/05/2022
									A070101	Failure to Capture	15/05/2022	15/05/2022
									A070102	High Capture Threshold	15/05/2022	15/05/2022
									A072201	High Impedance	15/05/2022	15/05/2022
							F1905	Device Revision or Replacement	A04	Material Integrity Problem	15/05/2022	15/05/2022
									A070101	Failure to Capture	15/05/2022	15/05/2022
									A070102	High Capture Threshold	15/05/2022	15/05/2022
									A072201	High Impedance	15/05/2022	15/05/2022
					CT3006	Implantable inorganic material	E2403	No Clinical Signs, Symptoms or Conditions	A04	Material Integrity Problem	15/05/2022	15/05/2022
									A070101	Failure to Capture	15/05/2022	15/05/2022
									A070102	High Capture Threshold	15/05/2022	15/05/2022
									A072201	High Impedance	15/05/2022	15/05/2022
							F08	Hospitalization or Prolonged Hospitalization	A04	Material Integrity Problem	15/05/2022	15/05/2022
									A070101	Failure to Capture	15/05/2022	15/05/2022
									A070102	High Capture Threshold	15/05/2022	15/05/2022
									A072201	High Impedance	15/05/2022	15/05/2022
							F1905	Device Revision or Replacement	A04	Material Integrity Problem	15/05/2022	15/05/2022
									A070101	Failure to Capture	15/05/2022	15/05/2022
									A070102	High Capture Threshold	15/05/2022	15/05/2022
									A072201	High Impedance	15/05/2022	15/05/2022
					CT314	Electrically powered	E2403	No Clinical Signs, Symptoms or Conditions	A04	Material Integrity Problem	15/05/2022	15/05/2022
									A070101	Failure to Capture	15/05/2022	15/05/2022

									A070102	High Capture Threshold	15/05/2022	15/05/2022
									A072201	High Impedance	15/05/2022	15/05/2022
							F08	Hospitalization or Prolonged Hospitalization	A04	Material Integrity Problem	15/05/2022	15/05/2022
									A070101	Failure to Capture	15/05/2022	15/05/2022
									A070102	High Capture Threshold	15/05/2022	15/05/2022
									A072201	High Impedance	15/05/2022	15/05/2022
							F1905	Device Revision or Replacement	A04	Material Integrity Problem	15/05/2022	15/05/2022
									A070101	Failure to Capture	15/05/2022	15/05/2022
									A070102	High Capture Threshold	15/05/2022	15/05/2022
									A072201	High Impedance	15/05/2022	15/05/2022
					CT321	Long-term surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A04	Material Integrity Problem	15/05/2022	15/05/2022
									A070101	Failure to Capture	15/05/2022	15/05/2022
									A070102	High Capture Threshold	15/05/2022	15/05/2022
									A072201	High Impedance	15/05/2022	15/05/2022
							F08	Hospitalization or Prolonged Hospitalization	A04	Material Integrity Problem	15/05/2022	15/05/2022
									A070101	Failure to Capture	15/05/2022	15/05/2022
									A070102	High Capture Threshold	15/05/2022	15/05/2022
									A072201	High Impedance	15/05/2022	15/05/2022
							F1905	Device Revision or Replacement	A04	Material Integrity Problem	15/05/2022	15/05/2022
									A070101	Failure to Capture	15/05/2022	15/05/2022
									A070102	High Capture Threshold	15/05/2022	15/05/2022
									A072201	High Impedance	15/05/2022	15/05/2022
					CT449	Generators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A04	Material Integrity Problem	15/05/2022	15/05/2022
									A070101	Failure to Capture	15/05/2022	15/05/2022
									A070102	High Capture Threshold	15/05/2022	15/05/2022
									A072201	High Impedance	15/05/2022	15/05/2022
							F08	Hospitalization or Prolonged Hospitalization	A04	Material Integrity Problem	15/05/2022	15/05/2022
									A070101	Failure to Capture	15/05/2022	15/05/2022
									A070102	High Capture Threshold	15/05/2022	15/05/2022
									A072201	High Impedance	15/05/2022	15/05/2022
							F1905	Device Revision or Replacement	A04	Material Integrity Problem	15/05/2022	15/05/2022
									A070101	Failure to Capture	15/05/2022	15/05/2022
									A070102	High Capture Threshold	15/05/2022	15/05/2022
									A072201	High Impedance	15/05/2022	15/05/2022
					CT981	Single use	E2403	No Clinical Signs, Symptoms or Conditions	A04	Material Integrity Problem	15/05/2022	15/05/2022
									A070101	Failure to Capture	15/05/2022	15/05/2022
									A070102	High Capture Threshold	15/05/2022	15/05/2022
									A072201	High Impedance	15/05/2022	15/05/2022
							F08	Hospitalization or Prolonged Hospitalization	A04	Material Integrity Problem	15/05/2022	15/05/2022
									A070101	Failure to Capture	15/05/2022	15/05/2022
									A070102	High Capture Threshold	15/05/2022	15/05/2022
									A072201	High Impedance	15/05/2022	15/05/2022
							F1905	Device Revision or Replacement	A04	Material Integrity Problem	15/05/2022	15/05/2022
									A070101	Failure to Capture	15/05/2022	15/05/2022
									A070102	High Capture Threshold	15/05/2022	15/05/2022
									A072201	High Impedance	15/05/2022	15/05/2022
					CT983	Surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A04	Material Integrity Problem	15/05/2022	15/05/2022
									A070101	Failure to Capture	15/05/2022	15/05/2022
									A070102	High Capture Threshold	15/05/2022	15/05/2022
									A072201	High Impedance	15/05/2022	15/05/2022
							F08	Hospitalization or Prolonged Hospitalization	A04	Material Integrity Problem	15/05/2022	15/05/2022
									A070101	Failure to Capture	15/05/2022	15/05/2022
									A070102	High Capture Threshold	15/05/2022	15/05/2022
									A072201	High Impedance	15/05/2022	15/05/2022
							F1905	Device Revision or Replacement	A04	Material Integrity Problem	15/05/2022	15/05/2022
									A070101	Failure to Capture	15/05/2022	15/05/2022
									A070102	High Capture Threshold	15/05/2022	15/05/2022
									A072201	High Impedance	15/05/2022	15/05/2022
6670	2021	All other reportable incidents	60910	Coronary venous pacing lead	CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070101	Failure to Capture	23/09/2021	23/09/2021
									A072201	High Impedance	23/09/2021	23/09/2021
							F26	No Health Consequences or Impact	A070101	Failure to Capture	23/09/2021	23/09/2021
									A072201	High Impedance	23/09/2021	23/09/2021
6671	2021	All other reportable incidents	17882	Physiologic-monitoring defibrillation system	CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A0801	Failure to Calibrate		
							F27	Problem Identified During Non-Clinical Procedure	A0801	Failure to Calibrate		
6672	2021	All other reportable incidents	47265	Dual-chamber implantable pacemaker, rate-responsive	CT141	Electrodes/Leads and associated devices	E060102	Atrial Fibrillation	A0712	Pacing Problem	07/07/2021	07/07/2021
							E2403	No Clinical Signs, Symptoms or Conditions	A0712	Pacing Problem	07/07/2021	07/07/2021
							F1905	Device Revision or Replacement	A0712	Pacing Problem	07/07/2021	07/07/2021
					CT145	Cardiovascular devices	E060102	Atrial Fibrillation	A0712	Pacing Problem	07/07/2021	07/07/2021
							E2403	No Clinical Signs, Symptoms or Conditions	A0712	Pacing Problem	07/07/2021	07/07/2021
							F1905	Device Revision or Replacement	A0712	Pacing Problem	07/07/2021	07/07/2021
					CT1684	Cardiac pacemakers and associated devices	E060102	Atrial Fibrillation	A0712	Pacing Problem	07/07/2021	07/07/2021
							E2403	No Clinical Signs, Symptoms or Conditions	A0712	Pacing Problem	07/07/2021	07/07/2021
							F1905	Device Revision or Replacement	A0712	Pacing Problem	07/07/2021	07/07/2021
					CT1685	Cardiac pacemakers	E060102	Atrial Fibrillation	A0712	Pacing Problem	07/07/2021	07/07/2021
							E2403	No Clinical Signs, Symptoms or Conditions	A0712	Pacing Problem	07/07/2021	07/07/2021
							F1905	Device Revision or Replacement	A0712	Pacing Problem	07/07/2021	07/07/2021
					CT233	Surgical	E060102	Atrial Fibrillation	A0712	Pacing Problem	07/07/2021	07/07/2021
							E2403	No Clinical Signs, Symptoms or Conditions	A0712	Pacing Problem	07/07/2021	07/07/2021
							F1905	Device Revision or Replacement	A0712	Pacing Problem	07/07/2021	07/07/2021
					CT2406	Implantable	E060102	Atrial Fibrillation	A0712	Pacing Problem	07/07/2021	07/07/2021
							E2403	No Clinical Signs, Symptoms or Conditions	A0712	Pacing Problem	07/07/2021	07/07/2021
							F1905	Device Revision or Replacement	A0712	Pacing Problem	07/07/2021	07/07/2021
					CT2407	Active implantable	E060102	Atrial Fibrillation	A0712	Pacing Problem	07/07/2021	07/07/2021
							E2403	No Clinical Signs, Symptoms or Conditions	A0712	Pacing Problem	07/07/2021	07/07/2021
							F1905	Device Revision or Replacement	A0712	Pacing Problem	07/07/2021	07/07/2021
					CT257	Implantable cardiac pacemakers	E060102	Atrial Fibrillation	A0712	Pacing Problem	07/07/2021	07/07/2021
							E2403	No Clinical Signs, Symptoms or Conditions	A0712	Pacing Problem	07/07/2021	07/07/2021
							F1905	Device Revision or Replacement	A0712	Pacing Problem	07/07/2021	07/07/2021

					CT2727	Active implant devices/ancillaries	E060102	Atrial Fibrillation	A0712	Pacing Problem	07/07/2021	07/07/2021
							E2403	No Clinical Signs, Symptoms or Conditions	A0712	Pacing Problem	07/07/2021	07/07/2021
							F1905	Device Revision or Replacement	A0712	Pacing Problem	07/07/2021	07/07/2021
					CT2750	Implantable cardiac electrical stimulators and associated devices	E060102	Atrial Fibrillation	A0712	Pacing Problem	07/07/2021	07/07/2021
							E2403	No Clinical Signs, Symptoms or Conditions	A0712	Pacing Problem	07/07/2021	07/07/2021
							F1905	Device Revision or Replacement	A0712	Pacing Problem	07/07/2021	07/07/2021
					CT2866	Implantable cardiac electrical stimulators	E060102	Atrial Fibrillation	A0712	Pacing Problem	07/07/2021	07/07/2021
							E2403	No Clinical Signs, Symptoms or Conditions	A0712	Pacing Problem	07/07/2021	07/07/2021
							F1905	Device Revision or Replacement	A0712	Pacing Problem	07/07/2021	07/07/2021
					CT2985	Active	E060102	Atrial Fibrillation	A0712	Pacing Problem	07/07/2021	07/07/2021
							E2403	No Clinical Signs, Symptoms or Conditions	A0712	Pacing Problem	07/07/2021	07/07/2021
							F1905	Device Revision or Replacement	A0712	Pacing Problem	07/07/2021	07/07/2021
					CT3001	Implantable metal	E060102	Atrial Fibrillation	A0712	Pacing Problem	07/07/2021	07/07/2021
							E2403	No Clinical Signs, Symptoms or Conditions	A0712	Pacing Problem	07/07/2021	07/07/2021
							F1905	Device Revision or Replacement	A0712	Pacing Problem	07/07/2021	07/07/2021
					CT3002	Implantable synthetic polymer	E060102	Atrial Fibrillation	A0712	Pacing Problem	07/07/2021	07/07/2021
							E2403	No Clinical Signs, Symptoms or Conditions	A0712	Pacing Problem	07/07/2021	07/07/2021
							F1905	Device Revision or Replacement	A0712	Pacing Problem	07/07/2021	07/07/2021
					CT3006	Implantable inorganic material	E060102	Atrial Fibrillation	A0712	Pacing Problem	07/07/2021	07/07/2021
							E2403	No Clinical Signs, Symptoms or Conditions	A0712	Pacing Problem	07/07/2021	07/07/2021
							F1905	Device Revision or Replacement	A0712	Pacing Problem	07/07/2021	07/07/2021
					CT314	Electrically powered	E060102	Atrial Fibrillation	A0712	Pacing Problem	07/07/2021	07/07/2021
							E2403	No Clinical Signs, Symptoms or Conditions	A0712	Pacing Problem	07/07/2021	07/07/2021
							F1905	Device Revision or Replacement	A0712	Pacing Problem	07/07/2021	07/07/2021
					CT321	Long-term surgical invasive	E060102	Atrial Fibrillation	A0712	Pacing Problem	07/07/2021	07/07/2021
							E2403	No Clinical Signs, Symptoms or Conditions	A0712	Pacing Problem	07/07/2021	07/07/2021
							F1905	Device Revision or Replacement	A0712	Pacing Problem	07/07/2021	07/07/2021
					CT449	Generators and associated devices	E060102	Atrial Fibrillation	A0712	Pacing Problem	07/07/2021	07/07/2021
							E2403	No Clinical Signs, Symptoms or Conditions	A0712	Pacing Problem	07/07/2021	07/07/2021
							F1905	Device Revision or Replacement	A0712	Pacing Problem	07/07/2021	07/07/2021
					CT981	Single use	E060102	Atrial Fibrillation	A0712	Pacing Problem	07/07/2021	07/07/2021
							E2403	No Clinical Signs, Symptoms or Conditions	A0712	Pacing Problem	07/07/2021	07/07/2021
							F1905	Device Revision or Replacement	A0712	Pacing Problem	07/07/2021	07/07/2021
					CT983	Surgical invasive	E060102	Atrial Fibrillation	A0712	Pacing Problem	07/07/2021	07/07/2021
							E2403	No Clinical Signs, Symptoms or Conditions	A0712	Pacing Problem	07/07/2021	07/07/2021
							F1905	Device Revision or Replacement	A0712	Pacing Problem	07/07/2021	07/07/2021
6673	2021	All other reportable incidents	47265	Dual-chamber implantable pacemaker, rate-responsive	CT141	Electrodes/Leads and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	01/09/2021	01/09/2021
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	01/09/2021	01/09/2021
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	01/09/2021	01/09/2021
					CT145	Cardiovascular devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	01/09/2021	01/09/2021
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	01/09/2021	01/09/2021
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	01/09/2021	01/09/2021
					CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	01/09/2021	01/09/2021
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	01/09/2021	01/09/2021
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	01/09/2021	01/09/2021
					CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	01/09/2021	01/09/2021
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	01/09/2021	01/09/2021
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	01/09/2021	01/09/2021
					CT233	Surgical	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	01/09/2021	01/09/2021
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	01/09/2021	01/09/2021
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	01/09/2021	01/09/2021
					CT2406	Implantable	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	01/09/2021	01/09/2021
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	01/09/2021	01/09/2021
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	01/09/2021	01/09/2021
					CT2407	Active implantable	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	01/09/2021	01/09/2021
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	01/09/2021	01/09/2021
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	01/09/2021	01/09/2021
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	01/09/2021	01/09/2021
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	01/09/2021	01/09/2021
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	01/09/2021	01/09/2021
					CT2727	Active implant devices/ancillaries	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	01/09/2021	01/09/2021
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	01/09/2021	01/09/2021
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	01/09/2021	01/09/2021
					CT2750	Implantable cardiac electrical stimulators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	01/09/2021	01/09/2021
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	01/09/2021	01/09/2021
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	01/09/2021	01/09/2021
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	01/09/2021	01/09/2021
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	01/09/2021	01/09/2021
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	01/09/2021	01/09/2021
					CT2985	Active	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	01/09/2021	01/09/2021
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	01/09/2021	01/09/2021
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	01/09/2021	01/09/2021
					CT3001	Implantable metal	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	01/09/2021	01/09/2021
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	01/09/2021	01/09/2021
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	01/09/2021	01/09/2021
					CT3002	Implantable synthetic polymer	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	01/09/2021	01/09/2021
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	01/09/2021	01/09/2021
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	01/09/2021	01/09/2021
					CT3006	Implantable inorganic material	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	01/09/2021	01/09/2021
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	01/09/2021	01/09/2021
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	01/09/2021	01/09/2021
					CT314	Electrically powered	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	01/09/2021	01/09/2021
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	01/09/2021	01/09/2021
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	01/09/2021	01/09/2021
					CT321	Long-term surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	01/09/2021	01/09/2021
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	01/09/2021	01/09/2021
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	01/09/2021	01/09/2021
					CT449	Generators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	01/09/2021	01/09/2021

							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	01/09/2021	01/09/2021
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	01/09/2021	01/09/2021
					CT981	Single use	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	01/09/2021	01/09/2021
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	01/09/2021	01/09/2021
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	01/09/2021	01/09/2021
					CT983	Surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	01/09/2021	01/09/2021
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	01/09/2021	01/09/2021
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	01/09/2021	01/09/2021
6674	2021	All other reportable incidents	37265	Dual-chamber implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A072201	High Impedance		
							F26	No Health Consequences or Impact	A072201	High Impedance		
6675	2021	All other reportable incidents	36241	Pacemaker/defibrillator lead	CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070909	Oversensing		
									A090801	Signal Artifact/Noise		
							F26	No Health Consequences or Impact	A070909	Oversensing		
									A090801	Signal Artifact/Noise		
6676	2022	All other reportable incidents	47270	Cardiac resynchronization therapy implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	11/05/2022	11/05/2022
							F1901	Additional Surgery	A070504	Premature Discharge of Battery	11/05/2022	11/05/2022
6677	2021	All other reportable incidents	35223	Endocardial/interventricular septal pacing lead	CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A0401	Break		
									A0701	Capturing Problem		
									A070909	Oversensing		
									A0712	Pacing Problem		
							F1901	Additional Surgery	A0401	Break		
									A0701	Capturing Problem		
									A070909	Oversensing		
									A0712	Pacing Problem		
6678	2021	All other reportable incidents	35223	Endocardial/interventricular septal pacing lead	CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A040101	Fracture		
									A070101	Failure to Capture		
									A072201	High Impedance		
									A090801	Signal Artifact/Noise		
							F1901	Additional Surgery	A040101	Fracture		
									A070101	Failure to Capture		
									A072201	High Impedance		
									A090801	Signal Artifact/Noise		
6679	2021	All other reportable incidents	35852	Single-chamber implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A070909	Oversensing		
							F26	No Health Consequences or Impact	A070909	Oversensing		
6680	2022	All other reportable incidents	35223	Endocardial/interventricular septal pacing lead	CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070101	Failure to Capture		
									A070102	High Capture Threshold		
									A072201	High Impedance		
							F26	No Health Consequences or Impact	A070101	Failure to Capture		
									A070102	High Capture Threshold		
									A072201	High Impedance		
6681	2022	All other reportable incidents	37265	Dual-chamber implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A072201	High Impedance		
									A090801	Signal Artifact/Noise		
							F26	No Health Consequences or Impact	A072201	High Impedance		
									A090801	Signal Artifact/Noise		
6682	2022	All other reportable incidents	36241	Pacemaker/defibrillator lead	CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A072201	High Impedance		
							F26	No Health Consequences or Impact	A072201	High Impedance		
6683	2022	All other reportable incidents	47265	Dual-chamber implantable pacemaker, rate-responsive	CT141	Electrodes/Leads and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	25/05/2022	25/05/2022
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	25/05/2022	25/05/2022
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	25/05/2022	25/05/2022
					CT145	Cardiovascular devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	25/05/2022	25/05/2022
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	25/05/2022	25/05/2022
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	25/05/2022	25/05/2022
					CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	25/05/2022	25/05/2022
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	25/05/2022	25/05/2022
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	25/05/2022	25/05/2022
					CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	25/05/2022	25/05/2022
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	25/05/2022	25/05/2022
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	25/05/2022	25/05/2022
					CT233	Surgical	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	25/05/2022	25/05/2022
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	25/05/2022	25/05/2022
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	25/05/2022	25/05/2022
					CT2406	Implantable	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	25/05/2022	25/05/2022
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	25/05/2022	25/05/2022
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	25/05/2022	25/05/2022
					CT2407	Active implantable	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	25/05/2022	25/05/2022
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	25/05/2022	25/05/2022
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	25/05/2022	25/05/2022
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	25/05/2022	25/05/2022
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	25/05/2022	25/05/2022
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	25/05/2022	25/05/2022
					CT2727	Active implant devices/ancillaries	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	25/05/2022	25/05/2022
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	25/05/2022	25/05/2022
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	25/05/2022	25/05/2022
					CT2750	Implantable cardiac electrical stimulators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	25/05/2022	25/05/2022
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	25/05/2022	25/05/2022
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	25/05/2022	25/05/2022
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	25/05/2022	25/05/2022
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	25/05/2022	25/05/2022
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	25/05/2022	25/05/2022
					CT2985	Active	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	25/05/2022	25/05/2022
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	25/05/2022	25/05/2022
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	25/05/2022	25/05/2022
					CT3001	Implantable metal	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	25/05/2022	25/05/2022
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	25/05/2022	25/05/2022
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	25/05/2022	25/05/2022
					CT3002	Implantable synthetic polymer	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	25/05/2022	25/05/2022
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	25/05/2022	25/05/2022
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	25/05/2022	25/05/2022

					CT3006	Implantable inorganic material	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	25/05/2022	25/05/2022
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	25/05/2022	25/05/2022
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	25/05/2022	25/05/2022
					CT314	Electrically powered	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	25/05/2022	25/05/2022
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	25/05/2022	25/05/2022
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	25/05/2022	25/05/2022
					CT321	Long-term surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	25/05/2022	25/05/2022
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	25/05/2022	25/05/2022
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	25/05/2022	25/05/2022
					CT449	Generators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	25/05/2022	25/05/2022
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	25/05/2022	25/05/2022
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	25/05/2022	25/05/2022
					CT981	Single use	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	25/05/2022	25/05/2022
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	25/05/2022	25/05/2022
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	25/05/2022	25/05/2022
					CT983	Surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	25/05/2022	25/05/2022
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	25/05/2022	25/05/2022
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	25/05/2022	25/05/2022
6684	2022	All other reportable incidents	47270	Cardiac resynchronization therapy implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	22/04/2022	22/04/2022
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	22/04/2022	22/04/2022
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	22/04/2022	22/04/2022
6685	2022	All other reportable incidents	17882	Physiologic-monitoring defibrillation system	CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A0719	Unexpected Shutdown		
							F26	No Health Consequences or Impact	A0719	Unexpected Shutdown		
6686	2022	All other reportable incidents	37265	Dual-chamber implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
6687	2021	All other reportable incidents	47265	Dual-chamber implantable pacemaker, rate-responsive	CT141	Electrodes/Leads and associated devices	E0717	Dyspnea	A090804	False Positive Result		
							E2403	No Clinical Signs, Symptoms or Conditions	A090804	False Positive Result		
							F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A090804	False Positive Result		
							F11	Non-Serious Injury/ Illness/ Impairment	A090804	False Positive Result		
					CT145	Cardiovascular devices	E0717	Dyspnea	A090804	False Positive Result		
							E2403	No Clinical Signs, Symptoms or Conditions	A090804	False Positive Result		
							F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A090804	False Positive Result		
							F11	Non-Serious Injury/ Illness/ Impairment	A090804	False Positive Result		
					CT1684	Cardiac pacemakers and associated devices	E0717	Dyspnea	A090804	False Positive Result		
							E2403	No Clinical Signs, Symptoms or Conditions	A090804	False Positive Result		
							F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A090804	False Positive Result		
							F11	Non-Serious Injury/ Illness/ Impairment	A090804	False Positive Result		
					CT1685	Cardiac pacemakers	E0717	Dyspnea	A090804	False Positive Result		
							E2403	No Clinical Signs, Symptoms or Conditions	A090804	False Positive Result		
							F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A090804	False Positive Result		
							F11	Non-Serious Injury/ Illness/ Impairment	A090804	False Positive Result		
					CT233	Surgical	E0717	Dyspnea	A090804	False Positive Result		
							E2403	No Clinical Signs, Symptoms or Conditions	A090804	False Positive Result		
							F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A090804	False Positive Result		
							F11	Non-Serious Injury/ Illness/ Impairment	A090804	False Positive Result		
					CT2406	Implantable	E0717	Dyspnea	A090804	False Positive Result		
							E2403	No Clinical Signs, Symptoms or Conditions	A090804	False Positive Result		
							F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A090804	False Positive Result		
							F11	Non-Serious Injury/ Illness/ Impairment	A090804	False Positive Result		
					CT2407	Active implantable	E0717	Dyspnea	A090804	False Positive Result		
							E2403	No Clinical Signs, Symptoms or Conditions	A090804	False Positive Result		
							F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A090804	False Positive Result		
							F11	Non-Serious Injury/ Illness/ Impairment	A090804	False Positive Result		
					CT257	Implantable cardiac pacemakers	E0717	Dyspnea	A090804	False Positive Result		
							E2403	No Clinical Signs, Symptoms or Conditions	A090804	False Positive Result		
							F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A090804	False Positive Result		
							F11	Non-Serious Injury/ Illness/ Impairment	A090804	False Positive Result		
					CT2727	Active implant devices/ancillaries	E0717	Dyspnea	A090804	False Positive Result		
							E2403	No Clinical Signs, Symptoms or Conditions	A090804	False Positive Result		
							F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A090804	False Positive Result		
							F11	Non-Serious Injury/ Illness/ Impairment	A090804	False Positive Result		
					CT2750	Implantable cardiac electrical stimulators and associated devices	E0717	Dyspnea	A090804	False Positive Result		
							E2403	No Clinical Signs, Symptoms or Conditions	A090804	False Positive Result		
							F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A090804	False Positive Result		
							F11	Non-Serious Injury/ Illness/ Impairment	A090804	False Positive Result		
					CT2866	Implantable cardiac electrical stimulators	E0717	Dyspnea	A090804	False Positive Result		
							E2403	No Clinical Signs, Symptoms or Conditions	A090804	False Positive Result		
							F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A090804	False Positive Result		
							F11	Non-Serious Injury/ Illness/ Impairment	A090804	False Positive Result		
					CT2985	Active	E0717	Dyspnea	A090804	False Positive Result		
							E2403	No Clinical Signs, Symptoms or Conditions	A090804	False Positive Result		
							F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A090804	False Positive Result		
							F11	Non-Serious Injury/ Illness/ Impairment	A090804	False Positive Result		
					CT3001	Implantable metal	E0717	Dyspnea	A090804	False Positive Result		
							E2403	No Clinical Signs, Symptoms or Conditions	A090804	False Positive Result		
							F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A090804	False Positive Result		
							F11	Non-Serious Injury/ Illness/ Impairment	A090804	False Positive Result		
					CT3002	Implantable synthetic polymer	E0717	Dyspnea	A090804	False Positive Result		
							E2403	No Clinical Signs, Symptoms or Conditions	A090804	False Positive Result		
							F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A090804	False Positive Result		
							F11	Non-Serious Injury/ Illness/ Impairment	A090804	False Positive Result		
					CT3006	Implantable inorganic material	E0717	Dyspnea	A090804	False Positive Result		
							E2403	No Clinical Signs, Symptoms or Conditions	A090804	False Positive Result		
							F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A090804	False Positive Result		
							F11	Non-Serious Injury/ Illness/ Impairment	A090804	False Positive Result		
					CT314	Electrically powered	E0717	Dyspnea	A090804	False Positive Result		

							E2403	No Clinical Signs, Symptoms or Conditions	A090804	False Positive Result		
							F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A090804	False Positive Result		
							F11	Non-Serious Injury/ Illness/ Impairment	A090804	False Positive Result		
					CT321	Long-term surgical invasive	E0717	Dyspnea	A090804	False Positive Result		
							E2403	No Clinical Signs, Symptoms or Conditions	A090804	False Positive Result		
							F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A090804	False Positive Result		
							F11	Non-Serious Injury/ Illness/ Impairment	A090804	False Positive Result		
					CT449	Generators and associated devices	E0717	Dyspnea	A090804	False Positive Result		
							E2403	No Clinical Signs, Symptoms or Conditions	A090804	False Positive Result		
							F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A090804	False Positive Result		
							F11	Non-Serious Injury/ Illness/ Impairment	A090804	False Positive Result		
					CT981	Single use	E0717	Dyspnea	A090804	False Positive Result		
							E2403	No Clinical Signs, Symptoms or Conditions	A090804	False Positive Result		
							F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A090804	False Positive Result		
							F11	Non-Serious Injury/ Illness/ Impairment	A090804	False Positive Result		
					CT983	Surgical invasive	E0717	Dyspnea	A090804	False Positive Result		
							E2403	No Clinical Signs, Symptoms or Conditions	A090804	False Positive Result		
							F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A090804	False Positive Result		
							F11	Non-Serious Injury/ Illness/ Impairment	A090804	False Positive Result		
6688	2021	All other reportable incidents	35852	Single-chamber implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem	14/10/2021	14/10/2021
									A070504	Premature Discharge of Battery	14/10/2021	14/10/2021
							F08	Hospitalization or Prolonged Hospitalization	A0705	Battery Problem	14/10/2021	14/10/2021
									A070504	Premature Discharge of Battery	14/10/2021	14/10/2021
							F1905	Device Revision or Replacement	A0705	Battery Problem	14/10/2021	14/10/2021
									A070504	Premature Discharge of Battery	14/10/2021	14/10/2021
6689	2021	All other reportable incidents	36241	Pacemaker/defibrillator lead	CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A0203	Defective Device	16/09/2021	16/09/2021
							F1908	Prolonged Surgery	A0203	Defective Device	16/09/2021	16/09/2021
6690	2022	All other reportable incidents	47270	Cardiac resynchronization therapy implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	18/05/2022	18/05/2022
							F1901	Additional Surgery	A070504	Premature Discharge of Battery	18/05/2022	18/05/2022
6691	2022	All other reportable incidents	37265	Dual-chamber implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E0601	Arrhythmia	A071302	Inappropriate/Inadequate Shock/Stimulation		
									A090805	Incorrect Measurement		
							E2403	No Clinical Signs, Symptoms or Conditions	A071302	Inappropriate/Inadequate Shock/Stimulation		
									A090805	Incorrect Measurement		
							F11	Non-Serious Injury/ Illness/ Impairment	A071302	Inappropriate/Inadequate Shock/Stimulation		
									A090805	Incorrect Measurement		
6692	2022	All other reportable incidents	47265	Dual-chamber implantable pacemaker, rate-responsive	CT141	Electrodes/Leads and associated devices	E011902	Presyncope	A071102	Failure to Interrogate		
									A071204	No Pacing		
							E2403	No Clinical Signs, Symptoms or Conditions	A071102	Failure to Interrogate		
									A071204	No Pacing		
							F26	No Health Consequences or Impact	A071102	Failure to Interrogate		
									A071204	No Pacing		
					CT145	Cardiovascular devices	E011902	Presyncope	A071102	Failure to Interrogate		
									A071204	No Pacing		
							E2403	No Clinical Signs, Symptoms or Conditions	A071102	Failure to Interrogate		
									A071204	No Pacing		
							F26	No Health Consequences or Impact	A071102	Failure to Interrogate		
									A071204	No Pacing		
					CT1684	Cardiac pacemakers and associated devices	E011902	Presyncope	A071102	Failure to Interrogate		
									A071204	No Pacing		
							E2403	No Clinical Signs, Symptoms or Conditions	A071102	Failure to Interrogate		
									A071204	No Pacing		
							F26	No Health Consequences or Impact	A071102	Failure to Interrogate		
									A071204	No Pacing		
					CT1685	Cardiac pacemakers	E011902	Presyncope	A071102	Failure to Interrogate		
									A071204	No Pacing		
							E2403	No Clinical Signs, Symptoms or Conditions	A071102	Failure to Interrogate		
									A071204	No Pacing		
							F26	No Health Consequences or Impact	A071102	Failure to Interrogate		
									A071204	No Pacing		
					CT233	Surgical	E011902	Presyncope	A071102	Failure to Interrogate		
									A071204	No Pacing		
							E2403	No Clinical Signs, Symptoms or Conditions	A071102	Failure to Interrogate		
									A071204	No Pacing		
							F26	No Health Consequences or Impact	A071102	Failure to Interrogate		
									A071204	No Pacing		
					CT2406	Implantable	E011902	Presyncope	A071102	Failure to Interrogate		
									A071204	No Pacing		
							E2403	No Clinical Signs, Symptoms or Conditions	A071102	Failure to Interrogate		
									A071204	No Pacing		
							F26	No Health Consequences or Impact	A071102	Failure to Interrogate		
									A071204	No Pacing		
					CT2407	Active implantable	E011902	Presyncope	A071102	Failure to Interrogate		
									A071204	No Pacing		
							E2403	No Clinical Signs, Symptoms or Conditions	A071102	Failure to Interrogate		
									A071204	No Pacing		
							F26	No Health Consequences or Impact	A071102	Failure to Interrogate		
									A071204	No Pacing		
					CT257	Implantable cardiac pacemakers	E011902	Presyncope	A071102	Failure to Interrogate		
									A071204	No Pacing		
							E2403	No Clinical Signs, Symptoms or Conditions	A071102	Failure to Interrogate		
									A071204	No Pacing		
							F26	No Health Consequences or Impact	A071102	Failure to Interrogate		
									A071204	No Pacing		
					CT2727	Active implant devices/ancillaries	E011902	Presyncope	A071102	Failure to Interrogate		
									A071204	No Pacing		
							E2403	No Clinical Signs, Symptoms or Conditions	A071102	Failure to Interrogate		
									A071204	No Pacing		
							F26	No Health Consequences or Impact	A071102	Failure to Interrogate		

									A071204	No Pacing		
					CT2750	Implantable cardiac electrical stimulators and associated devices	E011902	Presyncope	A071102	Failure to Interrogate		
									A071204	No Pacing		
							E2403	No Clinical Signs, Symptoms or Conditions	A071102	Failure to Interrogate		
									A071204	No Pacing		
							F26	No Health Consequences or Impact	A071102	Failure to Interrogate		
									A071204	No Pacing		
					CT2866	Implantable cardiac electrical stimulators	E011902	Presyncope	A071102	Failure to Interrogate		
									A071204	No Pacing		
							E2403	No Clinical Signs, Symptoms or Conditions	A071102	Failure to Interrogate		
									A071204	No Pacing		
							F26	No Health Consequences or Impact	A071102	Failure to Interrogate		
									A071204	No Pacing		
					CT2985	Active	E011902	Presyncope	A071102	Failure to Interrogate		
									A071204	No Pacing		
							E2403	No Clinical Signs, Symptoms or Conditions	A071102	Failure to Interrogate		
									A071204	No Pacing		
							F26	No Health Consequences or Impact	A071102	Failure to Interrogate		
									A071204	No Pacing		
					CT3001	Implantable metal	E011902	Presyncope	A071102	Failure to Interrogate		
									A071204	No Pacing		
							E2403	No Clinical Signs, Symptoms or Conditions	A071102	Failure to Interrogate		
									A071204	No Pacing		
							F26	No Health Consequences or Impact	A071102	Failure to Interrogate		
									A071204	No Pacing		
					CT3002	Implantable synthetic polymer	E011902	Presyncope	A071102	Failure to Interrogate		
									A071204	No Pacing		
							E2403	No Clinical Signs, Symptoms or Conditions	A071102	Failure to Interrogate		
									A071204	No Pacing		
							F26	No Health Consequences or Impact	A071102	Failure to Interrogate		
									A071204	No Pacing		
					CT3006	Implantable inorganic material	E011902	Presyncope	A071102	Failure to Interrogate		
									A071204	No Pacing		
							E2403	No Clinical Signs, Symptoms or Conditions	A071102	Failure to Interrogate		
									A071204	No Pacing		
							F26	No Health Consequences or Impact	A071102	Failure to Interrogate		
									A071204	No Pacing		
					CT314	Electrically powered	E011902	Presyncope	A071102	Failure to Interrogate		
									A071204	No Pacing		
							E2403	No Clinical Signs, Symptoms or Conditions	A071102	Failure to Interrogate		
									A071204	No Pacing		
							F26	No Health Consequences or Impact	A071102	Failure to Interrogate		
									A071204	No Pacing		
					CT321	Long-term surgical invasive	E011902	Presyncope	A071102	Failure to Interrogate		
									A071204	No Pacing		
							E2403	No Clinical Signs, Symptoms or Conditions	A071102	Failure to Interrogate		
									A071204	No Pacing		
							F26	No Health Consequences or Impact	A071102	Failure to Interrogate		
									A071204	No Pacing		
					CT449	Generators and associated devices	E011902	Presyncope	A071102	Failure to Interrogate		
									A071204	No Pacing		
							E2403	No Clinical Signs, Symptoms or Conditions	A071102	Failure to Interrogate		
									A071204	No Pacing		
							F26	No Health Consequences or Impact	A071102	Failure to Interrogate		
									A071204	No Pacing		
					CT981	Single use	E011902	Presyncope	A071102	Failure to Interrogate		
									A071204	No Pacing		
							E2403	No Clinical Signs, Symptoms or Conditions	A071102	Failure to Interrogate		
									A071204	No Pacing		
							F26	No Health Consequences or Impact	A071102	Failure to Interrogate		
									A071204	No Pacing		
					CT983	Surgical invasive	E011902	Presyncope	A071102	Failure to Interrogate		
									A071204	No Pacing		
							E2403	No Clinical Signs, Symptoms or Conditions	A071102	Failure to Interrogate		
									A071204	No Pacing		
							F26	No Health Consequences or Impact	A071102	Failure to Interrogate		
									A071204	No Pacing		
6693	2022	All other reportable incidents	17882	Physiologic-monitoring defibrillation system	CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A150105	Key or Button Unresponsive/not Working		
							F26	No Health Consequences or Impact	A150105	Key or Button Unresponsive/not Working		
6694	2022	All other reportable incidents	47267	Single-chamber implantable pacemaker, rate-responsive	CT141	Electrodes/Leads and associated devices	E060101	Asystole	A04	Material Integrity Problem	27/04/2022	27/04/2022
									A0712	Pacing Problem	27/04/2022	27/04/2022
									A072201	High Impedance	27/04/2022	27/04/2022
							E2403	No Clinical Signs, Symptoms or Conditions	A04	Material Integrity Problem	27/04/2022	27/04/2022
									A0712	Pacing Problem	27/04/2022	27/04/2022
									A072201	High Impedance	27/04/2022	27/04/2022
							F1001	Absence of Treatment	A04	Material Integrity Problem	27/04/2022	27/04/2022
									A0712	Pacing Problem	27/04/2022	27/04/2022
									A072201	High Impedance	27/04/2022	27/04/2022
							F1908	Prolonged Surgery	A04	Material Integrity Problem	27/04/2022	27/04/2022
									A0712	Pacing Problem	27/04/2022	27/04/2022
									A072201	High Impedance	27/04/2022	27/04/2022
					CT145	Cardiovascular devices	E060101	Asystole	A04	Material Integrity Problem	27/04/2022	27/04/2022
									A0712	Pacing Problem	27/04/2022	27/04/2022
									A072201	High Impedance	27/04/2022	27/04/2022
							E2403	No Clinical Signs, Symptoms or Conditions	A04	Material Integrity Problem	27/04/2022	27/04/2022
									A0712	Pacing Problem	27/04/2022	27/04/2022
									A072201	High Impedance	27/04/2022	27/04/2022
							F1001	Absence of Treatment	A04	Material Integrity Problem	27/04/2022	27/04/2022

								A0712	Pacing Problem	27/04/2022	27/04/2022
								A072201	High Impedance	27/04/2022	27/04/2022
							F1908	Prolonged Surgery	A04	Material Integrity Problem	27/04/2022	27/04/2022
									A0712	Pacing Problem	27/04/2022	27/04/2022
									A072201	High Impedance	27/04/2022	27/04/2022
					CT1684	Cardiac pacemakers and associated devices	E060101	Asystole	A04	Material Integrity Problem	27/04/2022	27/04/2022
									A0712	Pacing Problem	27/04/2022	27/04/2022
									A072201	High Impedance	27/04/2022	27/04/2022
							E2403	No Clinical Signs, Symptoms or Conditions	A04	Material Integrity Problem	27/04/2022	27/04/2022
									A0712	Pacing Problem	27/04/2022	27/04/2022
									A072201	High Impedance	27/04/2022	27/04/2022
							F1001	Absence of Treatment	A04	Material Integrity Problem	27/04/2022	27/04/2022
									A0712	Pacing Problem	27/04/2022	27/04/2022
									A072201	High Impedance	27/04/2022	27/04/2022
							F1908	Prolonged Surgery	A04	Material Integrity Problem	27/04/2022	27/04/2022
									A0712	Pacing Problem	27/04/2022	27/04/2022
									A072201	High Impedance	27/04/2022	27/04/2022
					CT1685	Cardiac pacemakers	E060101	Asystole	A04	Material Integrity Problem	27/04/2022	27/04/2022
									A0712	Pacing Problem	27/04/2022	27/04/2022
									A072201	High Impedance	27/04/2022	27/04/2022
							E2403	No Clinical Signs, Symptoms or Conditions	A04	Material Integrity Problem	27/04/2022	27/04/2022
									A0712	Pacing Problem	27/04/2022	27/04/2022
									A072201	High Impedance	27/04/2022	27/04/2022
							F1001	Absence of Treatment	A04	Material Integrity Problem	27/04/2022	27/04/2022
									A0712	Pacing Problem	27/04/2022	27/04/2022
									A072201	High Impedance	27/04/2022	27/04/2022
							F1908	Prolonged Surgery	A04	Material Integrity Problem	27/04/2022	27/04/2022
									A0712	Pacing Problem	27/04/2022	27/04/2022
									A072201	High Impedance	27/04/2022	27/04/2022
					CT233	Surgical	E060101	Asystole	A04	Material Integrity Problem	27/04/2022	27/04/2022
									A0712	Pacing Problem	27/04/2022	27/04/2022
									A072201	High Impedance	27/04/2022	27/04/2022
							E2403	No Clinical Signs, Symptoms or Conditions	A04	Material Integrity Problem	27/04/2022	27/04/2022
									A0712	Pacing Problem	27/04/2022	27/04/2022
									A072201	High Impedance	27/04/2022	27/04/2022
							F1001	Absence of Treatment	A04	Material Integrity Problem	27/04/2022	27/04/2022
									A0712	Pacing Problem	27/04/2022	27/04/2022
									A072201	High Impedance	27/04/2022	27/04/2022
							F1908	Prolonged Surgery	A04	Material Integrity Problem	27/04/2022	27/04/2022
									A0712	Pacing Problem	27/04/2022	27/04/2022
									A072201	High Impedance	27/04/2022	27/04/2022
					CT2406	Implantable	E060101	Asystole	A04	Material Integrity Problem	27/04/2022	27/04/2022
									A0712	Pacing Problem	27/04/2022	27/04/2022
									A072201	High Impedance	27/04/2022	27/04/2022
							E2403	No Clinical Signs, Symptoms or Conditions	A04	Material Integrity Problem	27/04/2022	27/04/2022
									A0712	Pacing Problem	27/04/2022	27/04/2022
									A072201	High Impedance	27/04/2022	27/04/2022
							F1001	Absence of Treatment	A04	Material Integrity Problem	27/04/2022	27/04/2022
									A0712	Pacing Problem	27/04/2022	27/04/2022
									A072201	High Impedance	27/04/2022	27/04/2022
							F1908	Prolonged Surgery	A04	Material Integrity Problem	27/04/2022	27/04/2022
									A0712	Pacing Problem	27/04/2022	27/04/2022
									A072201	High Impedance	27/04/2022	27/04/2022
					CT2407	Active implantable	E060101	Asystole	A04	Material Integrity Problem	27/04/2022	27/04/2022
									A0712	Pacing Problem	27/04/2022	27/04/2022
									A072201	High Impedance	27/04/2022	27/04/2022
							E2403	No Clinical Signs, Symptoms or Conditions	A04	Material Integrity Problem	27/04/2022	27/04/2022
									A0712	Pacing Problem	27/04/2022	27/04/2022
									A072201	High Impedance	27/04/2022	27/04/2022
							F1001	Absence of Treatment	A04	Material Integrity Problem	27/04/2022	27/04/2022
									A0712	Pacing Problem	27/04/2022	27/04/2022
									A072201	High Impedance	27/04/2022	27/04/2022
							F1908	Prolonged Surgery	A04	Material Integrity Problem	27/04/2022	27/04/2022
									A0712	Pacing Problem	27/04/2022	27/04/2022
									A072201	High Impedance	27/04/2022	27/04/2022
					CT257	Implantable cardiac pacemakers	E060101	Asystole	A04	Material Integrity Problem	27/04/2022	27/04/2022
									A0712	Pacing Problem	27/04/2022	27/04/2022
									A072201	High Impedance	27/04/2022	27/04/2022
							E2403	No Clinical Signs, Symptoms or Conditions	A04	Material Integrity Problem	27/04/2022	27/04/2022
									A0712	Pacing Problem	27/04/2022	27/04/2022
									A072201	High Impedance	27/04/2022	27/04/2022
							F1001	Absence of Treatment	A04	Material Integrity Problem	27/04/2022	27/04/2022
									A0712	Pacing Problem	27/04/2022	27/04/2022
									A072201	High Impedance	27/04/2022	27/04/2022
							F1908	Prolonged Surgery	A04	Material Integrity Problem	27/04/2022	27/04/2022
									A0712	Pacing Problem	27/04/2022	27/04/2022
									A072201	High Impedance	27/04/2022	27/04/2022
					CT2727	Active implant devices/ancillaries	E060101	Asystole	A04	Material Integrity Problem	27/04/2022	27/04/2022
									A0712	Pacing Problem	27/04/2022	27/04/2022
									A072201	High Impedance	27/04/2022	27/04/2022
							E2403	No Clinical Signs, Symptoms or Conditions	A04	Material Integrity Problem	27/04/2022	27/04/2022
									A0712	Pacing Problem	27/04/2022	27/04/2022
									A072201	High Impedance	27/04/2022	27/04/2022
							F1001	Absence of Treatment	A04	Material Integrity Problem	27/04/2022	27/04/2022
									A0712	Pacing Problem	27/04/2022	27/04/2022
									A072201	High Impedance	27/04/2022	27/04/2022
							F1908	Prolonged Surgery	A04	Material Integrity Problem	27/04/2022	27/04/2022
									A0712	Pacing Problem	27/04/2022	27/04/2022

								A072201	High Impedance	27/04/2022	27/04/2022
					CT2750	Implantable cardiac electrical stimulators and associated devices	E060101	Asystole	A04	Material Integrity Problem	27/04/2022	27/04/2022
									A0712	Pacing Problem	27/04/2022	27/04/2022
									A072201	High Impedance	27/04/2022	27/04/2022
							E2403	No Clinical Signs, Symptoms or Conditions	A04	Material Integrity Problem	27/04/2022	27/04/2022
									A0712	Pacing Problem	27/04/2022	27/04/2022
									A072201	High Impedance	27/04/2022	27/04/2022
							F1001	Absence of Treatment	A04	Material Integrity Problem	27/04/2022	27/04/2022
									A0712	Pacing Problem	27/04/2022	27/04/2022
									A072201	High Impedance	27/04/2022	27/04/2022
							F1908	Prolonged Surgery	A04	Material Integrity Problem	27/04/2022	27/04/2022
									A0712	Pacing Problem	27/04/2022	27/04/2022
									A072201	High Impedance	27/04/2022	27/04/2022
					CT2866	Implantable cardiac electrical stimulators	E060101	Asystole	A04	Material Integrity Problem	27/04/2022	27/04/2022
									A0712	Pacing Problem	27/04/2022	27/04/2022
									A072201	High Impedance	27/04/2022	27/04/2022
							E2403	No Clinical Signs, Symptoms or Conditions	A04	Material Integrity Problem	27/04/2022	27/04/2022
									A0712	Pacing Problem	27/04/2022	27/04/2022
									A072201	High Impedance	27/04/2022	27/04/2022
							F1001	Absence of Treatment	A04	Material Integrity Problem	27/04/2022	27/04/2022
									A0712	Pacing Problem	27/04/2022	27/04/2022
									A072201	High Impedance	27/04/2022	27/04/2022
							F1908	Prolonged Surgery	A04	Material Integrity Problem	27/04/2022	27/04/2022
									A0712	Pacing Problem	27/04/2022	27/04/2022
									A072201	High Impedance	27/04/2022	27/04/2022
					CT2985	Active	E060101	Asystole	A04	Material Integrity Problem	27/04/2022	27/04/2022
									A0712	Pacing Problem	27/04/2022	27/04/2022
									A072201	High Impedance	27/04/2022	27/04/2022
							E2403	No Clinical Signs, Symptoms or Conditions	A04	Material Integrity Problem	27/04/2022	27/04/2022
									A0712	Pacing Problem	27/04/2022	27/04/2022
									A072201	High Impedance	27/04/2022	27/04/2022
							F1001	Absence of Treatment	A04	Material Integrity Problem	27/04/2022	27/04/2022
									A0712	Pacing Problem	27/04/2022	27/04/2022
									A072201	High Impedance	27/04/2022	27/04/2022
							F1908	Prolonged Surgery	A04	Material Integrity Problem	27/04/2022	27/04/2022
									A0712	Pacing Problem	27/04/2022	27/04/2022
									A072201	High Impedance	27/04/2022	27/04/2022
					CT3001	Implantable metal	E060101	Asystole	A04	Material Integrity Problem	27/04/2022	27/04/2022
									A0712	Pacing Problem	27/04/2022	27/04/2022
									A072201	High Impedance	27/04/2022	27/04/2022
							E2403	No Clinical Signs, Symptoms or Conditions	A04	Material Integrity Problem	27/04/2022	27/04/2022
									A0712	Pacing Problem	27/04/2022	27/04/2022
									A072201	High Impedance	27/04/2022	27/04/2022
							F1001	Absence of Treatment	A04	Material Integrity Problem	27/04/2022	27/04/2022
									A0712	Pacing Problem	27/04/2022	27/04/2022
									A072201	High Impedance	27/04/2022	27/04/2022
							F1908	Prolonged Surgery	A04	Material Integrity Problem	27/04/2022	27/04/2022
									A0712	Pacing Problem	27/04/2022	27/04/2022
									A072201	High Impedance	27/04/2022	27/04/2022
					CT3002	Implantable synthetic polymer	E060101	Asystole	A04	Material Integrity Problem	27/04/2022	27/04/2022
									A0712	Pacing Problem	27/04/2022	27/04/2022
									A072201	High Impedance	27/04/2022	27/04/2022
							E2403	No Clinical Signs, Symptoms or Conditions	A04	Material Integrity Problem	27/04/2022	27/04/2022
									A0712	Pacing Problem	27/04/2022	27/04/2022
									A072201	High Impedance	27/04/2022	27/04/2022
							F1001	Absence of Treatment	A04	Material Integrity Problem	27/04/2022	27/04/2022
									A0712	Pacing Problem	27/04/2022	27/04/2022
									A072201	High Impedance	27/04/2022	27/04/2022
							F1908	Prolonged Surgery	A04	Material Integrity Problem	27/04/2022	27/04/2022
									A0712	Pacing Problem	27/04/2022	27/04/2022
									A072201	High Impedance	27/04/2022	27/04/2022
					CT3006	Implantable inorganic material	E060101	Asystole	A04	Material Integrity Problem	27/04/2022	27/04/2022
									A0712	Pacing Problem	27/04/2022	27/04/2022
									A072201	High Impedance	27/04/2022	27/04/2022
							E2403	No Clinical Signs, Symptoms or Conditions	A04	Material Integrity Problem	27/04/2022	27/04/2022
									A0712	Pacing Problem	27/04/2022	27/04/2022
									A072201	High Impedance	27/04/2022	27/04/2022
							F1001	Absence of Treatment	A04	Material Integrity Problem	27/04/2022	27/04/2022
									A0712	Pacing Problem	27/04/2022	27/04/2022
									A072201	High Impedance	27/04/2022	27/04/2022
							F1908	Prolonged Surgery	A04	Material Integrity Problem	27/04/2022	27/04/2022
									A0712	Pacing Problem	27/04/2022	27/04/2022
									A072201	High Impedance	27/04/2022	27/04/2022
					CT314	Electrically powered	E060101	Asystole	A04	Material Integrity Problem	27/04/2022	27/04/2022
									A0712	Pacing Problem	27/04/2022	27/04/2022
									A072201	High Impedance	27/04/2022	27/04/2022
							E2403	No Clinical Signs, Symptoms or Conditions	A04	Material Integrity Problem	27/04/2022	27/04/2022
									A0712	Pacing Problem	27/04/2022	27/04/2022
									A072201	High Impedance	27/04/2022	27/04/2022
							F1001	Absence of Treatment	A04	Material Integrity Problem	27/04/2022	27/04/2022
									A0712	Pacing Problem	27/04/2022	27/04/2022
									A072201	High Impedance	27/04/2022	27/04/2022
							F1908	Prolonged Surgery	A04	Material Integrity Problem	27/04/2022	27/04/2022
									A0712	Pacing Problem	27/04/2022	27/04/2022
									A072201	High Impedance	27/04/2022	27/04/2022
					CT321	Long-term surgical invasive	E060101	Asystole	A04	Material Integrity Problem	27/04/2022	27/04/2022
									A0712	Pacing Problem	27/04/2022	27/04/2022
									A072201	High Impedance	27/04/2022	27/04/2022

							E2403	No Clinical Signs, Symptoms or Conditions	A04	Material Integrity Problem	27/04/2022	27/04/2022
									A0712	Pacing Problem	27/04/2022	27/04/2022
									A072201	High Impedance	27/04/2022	27/04/2022
							F1001	Absence of Treatment	A04	Material Integrity Problem	27/04/2022	27/04/2022
									A0712	Pacing Problem	27/04/2022	27/04/2022
									A072201	High Impedance	27/04/2022	27/04/2022
							F1908	Prolonged Surgery	A04	Material Integrity Problem	27/04/2022	27/04/2022
									A0712	Pacing Problem	27/04/2022	27/04/2022
									A072201	High Impedance	27/04/2022	27/04/2022
					CT449	Generators and associated devices	E060101	Asystole	A04	Material Integrity Problem	27/04/2022	27/04/2022
									A0712	Pacing Problem	27/04/2022	27/04/2022
									A072201	High Impedance	27/04/2022	27/04/2022
							E2403	No Clinical Signs, Symptoms or Conditions	A04	Material Integrity Problem	27/04/2022	27/04/2022
									A0712	Pacing Problem	27/04/2022	27/04/2022
									A072201	High Impedance	27/04/2022	27/04/2022
							F1001	Absence of Treatment	A04	Material Integrity Problem	27/04/2022	27/04/2022
									A0712	Pacing Problem	27/04/2022	27/04/2022
									A072201	High Impedance	27/04/2022	27/04/2022
							F1908	Prolonged Surgery	A04	Material Integrity Problem	27/04/2022	27/04/2022
									A0712	Pacing Problem	27/04/2022	27/04/2022
									A072201	High Impedance	27/04/2022	27/04/2022
					CT981	Single use	E060101	Asystole	A04	Material Integrity Problem	27/04/2022	27/04/2022
									A0712	Pacing Problem	27/04/2022	27/04/2022
									A072201	High Impedance	27/04/2022	27/04/2022
							E2403	No Clinical Signs, Symptoms or Conditions	A04	Material Integrity Problem	27/04/2022	27/04/2022
									A0712	Pacing Problem	27/04/2022	27/04/2022
									A072201	High Impedance	27/04/2022	27/04/2022
							F1001	Absence of Treatment	A04	Material Integrity Problem	27/04/2022	27/04/2022
									A0712	Pacing Problem	27/04/2022	27/04/2022
									A072201	High Impedance	27/04/2022	27/04/2022
							F1908	Prolonged Surgery	A04	Material Integrity Problem	27/04/2022	27/04/2022
									A0712	Pacing Problem	27/04/2022	27/04/2022
									A072201	High Impedance	27/04/2022	27/04/2022
					CT983	Surgical invasive	E060101	Asystole	A04	Material Integrity Problem	27/04/2022	27/04/2022
									A0712	Pacing Problem	27/04/2022	27/04/2022
									A072201	High Impedance	27/04/2022	27/04/2022
							E2403	No Clinical Signs, Symptoms or Conditions	A04	Material Integrity Problem	27/04/2022	27/04/2022
									A0712	Pacing Problem	27/04/2022	27/04/2022
									A072201	High Impedance	27/04/2022	27/04/2022
							F1001	Absence of Treatment	A04	Material Integrity Problem	27/04/2022	27/04/2022
									A0712	Pacing Problem	27/04/2022	27/04/2022
									A072201	High Impedance	27/04/2022	27/04/2022
							F1908	Prolonged Surgery	A04	Material Integrity Problem	27/04/2022	27/04/2022
									A0712	Pacing Problem	27/04/2022	27/04/2022
									A072201	High Impedance	27/04/2022	27/04/2022
6695	2021	All other reportable incidents	37265	Dual-chamber implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	20/10/2021	20/10/2021
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	20/10/2021	20/10/2021
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	20/10/2021	20/10/2021
6696	2022	All other reportable incidents	35223	Endocardial/interventricular septal pacing lead	CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A0401	Break		
									A070909	Oversensing		
							F23	Unexpected Medical Intervention	A0401	Break		
									A070909	Oversensing		
6697	2022	All other reportable incidents	35223	Endocardial/interventricular septal pacing lead	CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070901	Decreased Sensitivity		
									A070910	Undersensing		
							F26	No Health Consequences or Impact	A070901	Decreased Sensitivity		
									A070910	Undersensing		
6698	2022	All other reportable incidents	35223	Endocardial/interventricular septal pacing lead	CT1684	Cardiac pacemakers and associated devices	E2402	Appropriate Clinical Signs, Symptoms and Conditions Term/Code not Available	A090801	Signal Artifact/Noise	12/05/2022	12/05/2022
							E2403	No Clinical Signs, Symptoms or Conditions	A090801	Signal Artifact/Noise	12/05/2022	12/05/2022
							F28	Appropriate Health Impact Term/Code not Available	A090801	Signal Artifact/Noise	12/05/2022	12/05/2022
6699	2022	All other reportable incidents	47265	Dual-chamber implantable pacemaker, rate-responsive	CT141	Electrodes/Leads and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
					CT145	Cardiovascular devices	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
					CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
					CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
					CT233	Surgical	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
					CT2406	Implantable	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
					CT2407	Active implantable	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		

					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
					CT2727	Active implant devices/ancillaries	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
					CT2750	Implantable cardiac electrical stimulators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
					CT2985	Active	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
					CT3001	Implantable metal	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
					CT3002	Implantable synthetic polymer	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
					CT3006	Implantable inorganic material	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
					CT314	Electrically powered	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
					CT321	Long-term surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
					CT449	Generators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
					CT981	Single use	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
					CT983	Surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
6700	2022	All other reportable incidents	47265	Dual-chamber implantable pacemaker, rate-responsive	CT141	Electrodes/Leads and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A072201	High Impedance		
							F26	No Health Consequences or Impact	A072201	High Impedance		
					CT145	Cardiovascular devices	E2403	No Clinical Signs, Symptoms or Conditions	A072201	High Impedance		
							F26	No Health Consequences or Impact	A072201	High Impedance		
					CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A072201	High Impedance		
							F26	No Health Consequences or Impact	A072201	High Impedance		
					CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A072201	High Impedance		
							F26	No Health Consequences or Impact	A072201	High Impedance		
					CT233	Surgical	E2403	No Clinical Signs, Symptoms or Conditions	A072201	High Impedance		
							F26	No Health Consequences or Impact	A072201	High Impedance		
					CT2406	Implantable	E2403	No Clinical Signs, Symptoms or Conditions	A072201	High Impedance		
							F26	No Health Consequences or Impact	A072201	High Impedance		
					CT2407	Active implantable	E2403	No Clinical Signs, Symptoms or Conditions	A072201	High Impedance		
							F26	No Health Consequences or Impact	A072201	High Impedance		
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A072201	High Impedance		
							F26	No Health Consequences or Impact	A072201	High Impedance		
					CT2727	Active implant devices/ancillaries	E2403	No Clinical Signs, Symptoms or Conditions	A072201	High Impedance		
							F26	No Health Consequences or Impact	A072201	High Impedance		
					CT2750	Implantable cardiac electrical stimulators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A072201	High Impedance		
							F26	No Health Consequences or Impact	A072201	High Impedance		
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A072201	High Impedance		
							F26	No Health Consequences or Impact	A072201	High Impedance		
					CT2985	Active	E2403	No Clinical Signs, Symptoms or Conditions	A072201	High Impedance		
							F26	No Health Consequences or Impact	A072201	High Impedance		
					CT3001	Implantable metal	E2403	No Clinical Signs, Symptoms or Conditions	A072201	High Impedance		
							F26	No Health Consequences or Impact	A072201	High Impedance		
					CT3002	Implantable synthetic polymer	E2403	No Clinical Signs, Symptoms or Conditions	A072201	High Impedance		
							F26	No Health Consequences or Impact	A072201	High Impedance		
					CT3006	Implantable inorganic material	E2403	No Clinical Signs, Symptoms or Conditions	A072201	High Impedance		
							F26	No Health Consequences or Impact	A072201	High Impedance		
					CT314	Electrically powered	E2403	No Clinical Signs, Symptoms or Conditions	A072201	High Impedance		
							F26	No Health Consequences or Impact	A072201	High Impedance		
					CT321	Long-term surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A072201	High Impedance		
							F26	No Health Consequences or Impact	A072201	High Impedance		
					CT449	Generators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A072201	High Impedance		
							F26	No Health Consequences or Impact	A072201	High Impedance		

					CT981	Single use	E2403	No Clinical Signs, Symptoms or Conditions	A072201	High Impedance		
							F26	No Health Consequences or Impact	A072201	High Impedance		
					CT983	Surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A072201	High Impedance		
							F26	No Health Consequences or Impact	A072201	High Impedance		
6701	2022	All other reportable incidents	47270	Cardiac resynchronization therapy implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	12/05/2022	12/05/2022
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	12/05/2022	12/05/2022
							F12	Serious Injury/ Illness/ Impairment	A070504	Premature Discharge of Battery	12/05/2022	12/05/2022
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	12/05/2022	12/05/2022
6702	2021	All other reportable incidents	35852	Single-chamber implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A070909	Oversensing		
							F26	No Health Consequences or Impact	A070909	Oversensing		
6703	2021	All other reportable incidents	35223	Endocardial/interventricular septal pacing lead	CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A0401	Break		
									A070909	Oversensing		
									A090801	Signal Artifact/Noise		
							F26	No Health Consequences or Impact	A0401	Break		
									A070909	Oversensing		
									A090801	Signal Artifact/Noise		
6704	2021	All other reportable incidents	45966	Transthoracic pacing lead	CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A0401	Break		
							F26	No Health Consequences or Impact	A0401	Break		
6705	2022	All other reportable incidents	47265	Dual-chamber implantable pacemaker, rate-responsive	CT141	Electrodes/Leads and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem	23/06/2022	23/06/2022
									A070504	Premature Discharge of Battery	23/06/2022	23/06/2022
							F08	Hospitalization or Prolonged Hospitalization	A0705	Battery Problem	23/06/2022	23/06/2022
									A070504	Premature Discharge of Battery	23/06/2022	23/06/2022
							F1905	Device Revision or Replacement	A0705	Battery Problem	23/06/2022	23/06/2022
									A070504	Premature Discharge of Battery	23/06/2022	23/06/2022
					CT145	Cardiovascular devices	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem	23/06/2022	23/06/2022
									A070504	Premature Discharge of Battery	23/06/2022	23/06/2022
							F08	Hospitalization or Prolonged Hospitalization	A0705	Battery Problem	23/06/2022	23/06/2022
									A070504	Premature Discharge of Battery	23/06/2022	23/06/2022
							F1905	Device Revision or Replacement	A0705	Battery Problem	23/06/2022	23/06/2022
									A070504	Premature Discharge of Battery	23/06/2022	23/06/2022
					CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem	23/06/2022	23/06/2022
									A070504	Premature Discharge of Battery	23/06/2022	23/06/2022
							F08	Hospitalization or Prolonged Hospitalization	A0705	Battery Problem	23/06/2022	23/06/2022
									A070504	Premature Discharge of Battery	23/06/2022	23/06/2022
							F1905	Device Revision or Replacement	A0705	Battery Problem	23/06/2022	23/06/2022
									A070504	Premature Discharge of Battery	23/06/2022	23/06/2022
					CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem	23/06/2022	23/06/2022
									A070504	Premature Discharge of Battery	23/06/2022	23/06/2022
							F08	Hospitalization or Prolonged Hospitalization	A0705	Battery Problem	23/06/2022	23/06/2022
									A070504	Premature Discharge of Battery	23/06/2022	23/06/2022
							F1905	Device Revision or Replacement	A0705	Battery Problem	23/06/2022	23/06/2022
									A070504	Premature Discharge of Battery	23/06/2022	23/06/2022
					CT233	Surgical	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem	23/06/2022	23/06/2022
									A070504	Premature Discharge of Battery	23/06/2022	23/06/2022
							F08	Hospitalization or Prolonged Hospitalization	A0705	Battery Problem	23/06/2022	23/06/2022
									A070504	Premature Discharge of Battery	23/06/2022	23/06/2022
							F1905	Device Revision or Replacement	A0705	Battery Problem	23/06/2022	23/06/2022
									A070504	Premature Discharge of Battery	23/06/2022	23/06/2022
					CT2406	Implantable	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem	23/06/2022	23/06/2022
									A070504	Premature Discharge of Battery	23/06/2022	23/06/2022
							F08	Hospitalization or Prolonged Hospitalization	A0705	Battery Problem	23/06/2022	23/06/2022
									A070504	Premature Discharge of Battery	23/06/2022	23/06/2022
							F1905	Device Revision or Replacement	A0705	Battery Problem	23/06/2022	23/06/2022
									A070504	Premature Discharge of Battery	23/06/2022	23/06/2022
					CT2407	Active implantable	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem	23/06/2022	23/06/2022
									A070504	Premature Discharge of Battery	23/06/2022	23/06/2022
							F08	Hospitalization or Prolonged Hospitalization	A0705	Battery Problem	23/06/2022	23/06/2022
									A070504	Premature Discharge of Battery	23/06/2022	23/06/2022
							F1905	Device Revision or Replacement	A0705	Battery Problem	23/06/2022	23/06/2022
									A070504	Premature Discharge of Battery	23/06/2022	23/06/2022
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem	23/06/2022	23/06/2022
									A070504	Premature Discharge of Battery	23/06/2022	23/06/2022
							F08	Hospitalization or Prolonged Hospitalization	A0705	Battery Problem	23/06/2022	23/06/2022
									A070504	Premature Discharge of Battery	23/06/2022	23/06/2022
							F1905	Device Revision or Replacement	A0705	Battery Problem	23/06/2022	23/06/2022
									A070504	Premature Discharge of Battery	23/06/2022	23/06/2022
					CT2727	Active implant devices/ancillaries	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem	23/06/2022	23/06/2022
									A070504	Premature Discharge of Battery	23/06/2022	23/06/2022
							F08	Hospitalization or Prolonged Hospitalization	A0705	Battery Problem	23/06/2022	23/06/2022
									A070504	Premature Discharge of Battery	23/06/2022	23/06/2022
							F1905	Device Revision or Replacement	A0705	Battery Problem	23/06/2022	23/06/2022
									A070504	Premature Discharge of Battery	23/06/2022	23/06/2022
					CT2750	Implantable cardiac electrical stimulators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem	23/06/2022	23/06/2022
									A070504	Premature Discharge of Battery	23/06/2022	23/06/2022
							F08	Hospitalization or Prolonged Hospitalization	A0705	Battery Problem	23/06/2022	23/06/2022
									A070504	Premature Discharge of Battery	23/06/2022	23/06/2022
							F1905	Device Revision or Replacement	A0705	Battery Problem	23/06/2022	23/06/2022
									A070504	Premature Discharge of Battery	23/06/2022	23/06/2022
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem	23/06/2022	23/06/2022
									A070504	Premature Discharge of Battery	23/06/2022	23/06/2022
							F08	Hospitalization or Prolonged Hospitalization	A0705	Battery Problem	23/06/2022	23/06/2022
									A070504	Premature Discharge of Battery	23/06/2022	23/06/2022
							F1905	Device Revision or Replacement	A0705	Battery Problem	23/06/2022	23/06/2022
									A070504	Premature Discharge of Battery	23/06/2022	23/06/2022
					CT2985	Active	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem	23/06/2022	23/06/2022
									A070504	Premature Discharge of Battery	23/06/2022	23/06/2022
							F08	Hospitalization or Prolonged Hospitalization	A0705	Battery Problem	23/06/2022	23/06/2022
									A070504	Premature Discharge of Battery	23/06/2022	23/06/2022

							F1905	Device Revision or Replacement	A0705	Battery Problem	23/06/2022	23/06/2022
									A070504	Premature Discharge of Battery	23/06/2022	23/06/2022
					CT3001	Implantable metal	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem	23/06/2022	23/06/2022
									A070504	Premature Discharge of Battery	23/06/2022	23/06/2022
							F08	Hospitalization or Prolonged Hospitalization	A0705	Battery Problem	23/06/2022	23/06/2022
									A070504	Premature Discharge of Battery	23/06/2022	23/06/2022
							F1905	Device Revision or Replacement	A0705	Battery Problem	23/06/2022	23/06/2022
									A070504	Premature Discharge of Battery	23/06/2022	23/06/2022
					CT3002	Implantable synthetic polymer	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem	23/06/2022	23/06/2022
									A070504	Premature Discharge of Battery	23/06/2022	23/06/2022
							F08	Hospitalization or Prolonged Hospitalization	A0705	Battery Problem	23/06/2022	23/06/2022
									A070504	Premature Discharge of Battery	23/06/2022	23/06/2022
							F1905	Device Revision or Replacement	A0705	Battery Problem	23/06/2022	23/06/2022
									A070504	Premature Discharge of Battery	23/06/2022	23/06/2022
					CT3006	Implantable inorganic material	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem	23/06/2022	23/06/2022
									A070504	Premature Discharge of Battery	23/06/2022	23/06/2022
							F08	Hospitalization or Prolonged Hospitalization	A0705	Battery Problem	23/06/2022	23/06/2022
									A070504	Premature Discharge of Battery	23/06/2022	23/06/2022
							F1905	Device Revision or Replacement	A0705	Battery Problem	23/06/2022	23/06/2022
									A070504	Premature Discharge of Battery	23/06/2022	23/06/2022
					CT314	Electrically powered	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem	23/06/2022	23/06/2022
									A070504	Premature Discharge of Battery	23/06/2022	23/06/2022
							F08	Hospitalization or Prolonged Hospitalization	A0705	Battery Problem	23/06/2022	23/06/2022
									A070504	Premature Discharge of Battery	23/06/2022	23/06/2022
							F1905	Device Revision or Replacement	A0705	Battery Problem	23/06/2022	23/06/2022
									A070504	Premature Discharge of Battery	23/06/2022	23/06/2022
					CT321	Long-term surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem	23/06/2022	23/06/2022
									A070504	Premature Discharge of Battery	23/06/2022	23/06/2022
							F08	Hospitalization or Prolonged Hospitalization	A0705	Battery Problem	23/06/2022	23/06/2022
									A070504	Premature Discharge of Battery	23/06/2022	23/06/2022
							F1905	Device Revision or Replacement	A0705	Battery Problem	23/06/2022	23/06/2022
									A070504	Premature Discharge of Battery	23/06/2022	23/06/2022
					CT449	Generators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem	23/06/2022	23/06/2022
									A070504	Premature Discharge of Battery	23/06/2022	23/06/2022
							F08	Hospitalization or Prolonged Hospitalization	A0705	Battery Problem	23/06/2022	23/06/2022
									A070504	Premature Discharge of Battery	23/06/2022	23/06/2022
							F1905	Device Revision or Replacement	A0705	Battery Problem	23/06/2022	23/06/2022
									A070504	Premature Discharge of Battery	23/06/2022	23/06/2022
					CT981	Single use	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem	23/06/2022	23/06/2022
									A070504	Premature Discharge of Battery	23/06/2022	23/06/2022
							F08	Hospitalization or Prolonged Hospitalization	A0705	Battery Problem	23/06/2022	23/06/2022
									A070504	Premature Discharge of Battery	23/06/2022	23/06/2022
							F1905	Device Revision or Replacement	A0705	Battery Problem	23/06/2022	23/06/2022
									A070504	Premature Discharge of Battery	23/06/2022	23/06/2022
					CT983	Surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem	23/06/2022	23/06/2022
									A070504	Premature Discharge of Battery	23/06/2022	23/06/2022
							F08	Hospitalization or Prolonged Hospitalization	A0705	Battery Problem	23/06/2022	23/06/2022
									A070504	Premature Discharge of Battery	23/06/2022	23/06/2022
							F1905	Device Revision or Replacement	A0705	Battery Problem	23/06/2022	23/06/2022
									A070504	Premature Discharge of Battery	23/06/2022	23/06/2022
6706	2021	All other reportable incidents	45966	Transthoracic pacing lead	CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A0401	Break		
							F26	No Health Consequences or Impact	A0401	Break		
6707	2021	All other reportable incidents	45966	Transthoracic pacing lead	CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A0401	Break		
							F26	No Health Consequences or Impact	A0401	Break		
6708	2021	All other reportable incidents	36241	Pacemaker/defibrillator lead	CT1684	Cardiac pacemakers and associated devices	E2104	Electric Shock	A072201	High Impedance		
							E2403	No Clinical Signs, Symptoms or Conditions	A072201	High Impedance		
							F2203	Imaging Required	A072201	High Impedance		
							F23	Unexpected Medical Intervention	A072201	High Impedance		
6709	2021	All other reportable incidents	17882	Physiologic-monitoring defibrillation system	CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A0708	Power Problem		
							F27	Problem Identified During Non-Clinical Procedure	A0708	Power Problem		
6710	2021	All other reportable incidents	36241	Pacemaker/defibrillator lead	CT1684	Cardiac pacemakers and associated devices	E050303	Pulmonary Embolism	A24	No Apparent Device Problem		
							E0604	Cardiac Perforation	A24	No Apparent Device Problem		
							E0605	Cardiac Tamponade	A24	No Apparent Device Problem		
							E0619	Pericardial Effusion	A24	No Apparent Device Problem		
							E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem		
							F08	Hospitalization or Prolonged Hospitalization	A24	No Apparent Device Problem		
							F1904	Device Repositioning	A24	No Apparent Device Problem		
6711	2021	All other reportable incidents	17882	Physiologic-monitoring defibrillation system	CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070503	Failure to Run on Battery		
							F26	No Health Consequences or Impact	A070503	Failure to Run on Battery		
6712	2022	All other reportable incidents	17882	Physiologic-monitoring defibrillation system	CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A090805	Incorrect Measurement		
							F27	Problem Identified During Non-Clinical Procedure	A090805	Incorrect Measurement		
6713	2021	All other reportable incidents	17882	Physiologic-monitoring defibrillation system	CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070503	Failure to Run on Battery		
							F27	Problem Identified During Non-Clinical Procedure	A070503	Failure to Run on Battery		
6714	2021	All other reportable incidents	17882	Physiologic-monitoring defibrillation system	CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A090102	Inaudible or Unclear Audible Prompt/Feedback		
							F27	Problem Identified During Non-Clinical Procedure	A090102	Inaudible or Unclear Audible Prompt/Feedback		
6715	2022	All other reportable incidents	36241	Pacemaker/defibrillator lead	CT1684	Cardiac pacemakers and associated devices	E2104	Electric Shock	A040101	Fracture	12/05/2022	12/05/2022
									A051201	Device Dislodged or Dislocated	12/05/2022	12/05/2022
									A070909	Oversensing	12/05/2022	12/05/2022
									A071302	Inappropriate/Inadequate Shock/Stimulation	12/05/2022	12/05/2022
							E2403	No Clinical Signs, Symptoms or Conditions	A040101	Fracture	12/05/2022	12/05/2022
									A051201	Device Dislodged or Dislocated	12/05/2022	12/05/2022
									A070909	Oversensing	12/05/2022	12/05/2022
									A071302	Inappropriate/Inadequate Shock/Stimulation	12/05/2022	12/05/2022
							F08	Hospitalization or Prolonged Hospitalization	A040101	Fracture	12/05/2022	12/05/2022
									A051201	Device Dislodged or Dislocated	12/05/2022	12/05/2022
									A070909	Oversensing	12/05/2022	12/05/2022
									A071302	Inappropriate/Inadequate Shock/Stimulation	12/05/2022	12/05/2022
							F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A040101	Fracture	12/05/2022	12/05/2022

									A051201	Device Dislodged or Dislocated	12/05/2022	12/05/2022
									A070909	Oversensing	12/05/2022	12/05/2022
									A071302	Inappropriate/Inadequate Shock/Stimulation	12/05/2022	12/05/2022
							F1905	Device Revision or Replacement	A040101	Fracture	12/05/2022	12/05/2022
									A051201	Device Dislodged or Dislocated	12/05/2022	12/05/2022
									A070909	Oversensing	12/05/2022	12/05/2022
									A071302	Inappropriate/Inadequate Shock/Stimulation	12/05/2022	12/05/2022
							F2203	Imaging Required	A040101	Fracture	12/05/2022	12/05/2022
									A051201	Device Dislodged or Dislocated	12/05/2022	12/05/2022
									A070909	Oversensing	12/05/2022	12/05/2022
									A071302	Inappropriate/Inadequate Shock/Stimulation	12/05/2022	12/05/2022
6716	2022	All other reportable incidents	47265	Dual-chamber implantable pacemaker, rate-responsive	CT141	Electrodes/Leads and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT145	Cardiovascular devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT233	Surgical	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT2406	Implantable	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT2407	Active implantable	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT2727	Active implant devices/ancillaries	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT2750	Implantable cardiac electrical stimulators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT2985	Active	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT3001	Implantable metal	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT3002	Implantable synthetic polymer	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT3006	Implantable inorganic material	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT314	Electrically powered	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT321	Long-term surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT449	Generators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT981	Single use	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT983	Surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
6717	2021	All other reportable incidents	35223	Endocardial/interventricular septal pacing lead	CT1684	Cardiac pacemakers and associated devices	E011903	Syncope/Fainting	A070101	Failure to Capture		
							E2403	No Clinical Signs, Symptoms or Conditions	A070101	Failure to Capture		
							F12	Serious Injury/ Illness/ Impairment	A070101	Failure to Capture		
6718	2021	All other reportable incidents	60789	Intracardiac pacemaker	CT130	Catheters and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A1502	Positioning Problem		
									A23	Use of Device Problem		
							F26	No Health Consequences or Impact	A1502	Positioning Problem		
									A23	Use of Device Problem		
					CT145	Cardiovascular devices	E2403	No Clinical Signs, Symptoms or Conditions	A1502	Positioning Problem		
									A23	Use of Device Problem		
							F26	No Health Consequences or Impact	A1502	Positioning Problem		
									A23	Use of Device Problem		
					CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A1502	Positioning Problem		
									A23	Use of Device Problem		
							F26	No Health Consequences or Impact	A1502	Positioning Problem		
									A23	Use of Device Problem		
					CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A1502	Positioning Problem		
									A23	Use of Device Problem		
							F26	No Health Consequences or Impact	A1502	Positioning Problem		
									A23	Use of Device Problem		
					CT233	Surgical	E2403	No Clinical Signs, Symptoms or Conditions	A1502	Positioning Problem		
									A23	Use of Device Problem		
							F26	No Health Consequences or Impact	A1502	Positioning Problem		
									A23	Use of Device Problem		
					CT2406	Implantable	E2403	No Clinical Signs, Symptoms or Conditions	A1502	Positioning Problem		
									A23	Use of Device Problem		
							F26	No Health Consequences or Impact	A1502	Positioning Problem		
									A23	Use of Device Problem		
					CT2407	Active implantable	E2403	No Clinical Signs, Symptoms or Conditions	A1502	Positioning Problem		
									A23	Use of Device Problem		
							F26	No Health Consequences or Impact	A1502	Positioning Problem		
									A23	Use of Device Problem		
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A1502	Positioning Problem		
									A23	Use of Device Problem		
							F26	No Health Consequences or Impact	A1502	Positioning Problem		
									A23	Use of Device Problem		
					CT2727	Active implant devices/ancillaries	E2403	No Clinical Signs, Symptoms or Conditions	A1502	Positioning Problem		
									A23	Use of Device Problem		

							F26	No Health Consequences or Impact	A1502	Positioning Problem		
									A23	Use of Device Problem		
					CT2750	Implantable cardiac electrical stimulators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A1502	Positioning Problem		
									A23	Use of Device Problem		
							F26	No Health Consequences or Impact	A1502	Positioning Problem		
									A23	Use of Device Problem		
					CT2818	Catheterization/Cannulation/Fluid path devices	E2403	No Clinical Signs, Symptoms or Conditions	A1502	Positioning Problem		
									A23	Use of Device Problem		
							F26	No Health Consequences or Impact	A1502	Positioning Problem		
									A23	Use of Device Problem		
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A1502	Positioning Problem		
									A23	Use of Device Problem		
							F26	No Health Consequences or Impact	A1502	Positioning Problem		
									A23	Use of Device Problem		
					CT2985	Active	E2403	No Clinical Signs, Symptoms or Conditions	A1502	Positioning Problem		
									A23	Use of Device Problem		
							F26	No Health Consequences or Impact	A1502	Positioning Problem		
									A23	Use of Device Problem		
					CT3001	Implantable metal	E2403	No Clinical Signs, Symptoms or Conditions	A1502	Positioning Problem		
									A23	Use of Device Problem		
							F26	No Health Consequences or Impact	A1502	Positioning Problem		
									A23	Use of Device Problem		
					CT3002	Implantable synthetic polymer	E2403	No Clinical Signs, Symptoms or Conditions	A1502	Positioning Problem		
									A23	Use of Device Problem		
							F26	No Health Consequences or Impact	A1502	Positioning Problem		
									A23	Use of Device Problem		
					CT3006	Implantable inorganic material	E2403	No Clinical Signs, Symptoms or Conditions	A1502	Positioning Problem		
									A23	Use of Device Problem		
							F26	No Health Consequences or Impact	A1502	Positioning Problem		
									A23	Use of Device Problem		
					CT314	Electrically powered	E2403	No Clinical Signs, Symptoms or Conditions	A1502	Positioning Problem		
									A23	Use of Device Problem		
							F26	No Health Consequences or Impact	A1502	Positioning Problem		
									A23	Use of Device Problem		
					CT321	Long-term surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A1502	Positioning Problem		
									A23	Use of Device Problem		
							F26	No Health Consequences or Impact	A1502	Positioning Problem		
									A23	Use of Device Problem		
					CT981	Single use	E2403	No Clinical Signs, Symptoms or Conditions	A1502	Positioning Problem		
									A23	Use of Device Problem		
							F26	No Health Consequences or Impact	A1502	Positioning Problem		
									A23	Use of Device Problem		
					CT983	Surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A1502	Positioning Problem		
									A23	Use of Device Problem		
							F26	No Health Consequences or Impact	A1502	Positioning Problem		
									A23	Use of Device Problem		
6719	2021	All other reportable incidents	36241	Pacemaker/defibrillator lead	CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A072202	Low Impedance		
							F26	No Health Consequences or Impact	A072202	Low Impedance		
6720	2021	All other reportable incidents	35223	Endocardial/interventricular septal pacing lead	CT1684	Cardiac pacemakers and associated devices	E0306	Sepsis	A070101	Failure to Capture		
									A070102	High Capture Threshold		
									A0709	Device Sensing Problem		
									A0712	Pacing Problem		
									A0722	Impedance Problem		
							E0604	Cardiac Perforation	A070101	Failure to Capture		
									A070102	High Capture Threshold		
									A0709	Device Sensing Problem		
									A0712	Pacing Problem		
									A0722	Impedance Problem		
							E0605	Cardiac Tamponade	A070101	Failure to Capture		
									A070102	High Capture Threshold		
									A0709	Device Sensing Problem		
									A0712	Pacing Problem		
									A0722	Impedance Problem		
							E0733	Pneumonia	A070101	Failure to Capture		
									A070102	High Capture Threshold		
									A0709	Device Sensing Problem		
									A0712	Pacing Problem		
									A0722	Impedance Problem		
							E1022	Perforation of Esophagus	A070101	Failure to Capture		
									A070102	High Capture Threshold		
									A0709	Device Sensing Problem		
									A0712	Pacing Problem		
									A0722	Impedance Problem		
							E1906	Unspecified Infection	A070101	Failure to Capture		
									A070102	High Capture Threshold		
									A0709	Device Sensing Problem		
									A0712	Pacing Problem		
									A0722	Impedance Problem		
							E2403	No Clinical Signs, Symptoms or Conditions	A070101	Failure to Capture		
									A070102	High Capture Threshold		
									A0709	Device Sensing Problem		
									A0712	Pacing Problem		
									A0722	Impedance Problem		
							F1203	Life Threatening Illness or Injury	A070101	Failure to Capture		
									A070102	High Capture Threshold		
									A0709	Device Sensing Problem		
									A0712	Pacing Problem		
									A0722	Impedance Problem		

							F1904	Device Repositioning	A070101	Failure to Capture		
									A070102	High Capture Threshold		
									A0709	Device Sensing Problem		
									A0712	Pacing Problem		
									A0722	Impedance Problem		
							F1905	Device Revision or Replacement	A070101	Failure to Capture		
									A070102	High Capture Threshold		
									A0709	Device Sensing Problem		
									A0712	Pacing Problem		
									A0722	Impedance Problem		
							F2203	Imaging Required	A070101	Failure to Capture		
									A070102	High Capture Threshold		
									A0709	Device Sensing Problem		
									A0712	Pacing Problem		
									A0722	Impedance Problem		
							F2302	Blood Transfusion	A070101	Failure to Capture		
									A070102	High Capture Threshold		
									A0709	Device Sensing Problem		
									A0712	Pacing Problem		
									A0722	Impedance Problem		
6721	2022	All other reportable incidents	17882	Physiologic-monitoring defibrillation system	CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem		
							F27	Problem Identified During Non-Clinical Procedure	A0705	Battery Problem		
6722	2021	All other reportable incidents	47265	Dual-chamber implantable pacemaker, rate-responsive	CT141	Electrodes/Leads and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT145	Cardiovascular devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT233	Surgical	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT2406	Implantable	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT2407	Active implantable	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT2727	Active implant devices/ancillaries	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT2750	Implantable cardiac electrical stimulators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT2985	Active	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT3001	Implantable metal	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT3002	Implantable synthetic polymer	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT3006	Implantable inorganic material	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT314	Electrically powered	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT321	Long-term surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT449	Generators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT981	Single use	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT983	Surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
6723	2021	All other reportable incidents	37265	Dual-chamber implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
6724	2021	All other reportable incidents	35852	Single-chamber implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A071101	Difficult to Interrogate		
							F26	No Health Consequences or Impact	A071101	Difficult to Interrogate		
6725	2021	All other reportable incidents	35852	Single-chamber implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem	13/10/2021	13/10/2021
									A070504	Premature Discharge of Battery	13/10/2021	13/10/2021
							F08	Hospitalization or Prolonged Hospitalization	A0705	Battery Problem	13/10/2021	13/10/2021
									A070504	Premature Discharge of Battery	13/10/2021	13/10/2021
							F1905	Device Revision or Replacement	A0705	Battery Problem	13/10/2021	13/10/2021
									A070504	Premature Discharge of Battery	13/10/2021	13/10/2021
6726	2022	All other reportable incidents	37265	Dual-chamber implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	18/05/2022	18/05/2022
							F1901	Additional Surgery	A070504	Premature Discharge of Battery	18/05/2022	18/05/2022
6727	2022	All other reportable incidents	47265	Dual-chamber implantable pacemaker, rate-responsive	CT141	Electrodes/Leads and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070909	Oversensing	16/05/2022	16/05/2022
							F1901	Additional Surgery	A070909	Oversensing	16/05/2022	16/05/2022
					CT145	Cardiovascular devices	E2403	No Clinical Signs, Symptoms or Conditions	A070909	Oversensing	16/05/2022	16/05/2022
							F1901	Additional Surgery	A070909	Oversensing	16/05/2022	16/05/2022
					CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070909	Oversensing	16/05/2022	16/05/2022
							F1901	Additional Surgery	A070909	Oversensing	16/05/2022	16/05/2022
					CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A070909	Oversensing	16/05/2022	16/05/2022
							F1901	Additional Surgery	A070909	Oversensing	16/05/2022	16/05/2022
					CT233	Surgical	E2403	No Clinical Signs, Symptoms or Conditions	A070909	Oversensing	16/05/2022	16/05/2022
							F1901	Additional Surgery	A070909	Oversensing	16/05/2022	16/05/2022
					CT2406	Implantable	E2403	No Clinical Signs, Symptoms or Conditions	A070909	Oversensing	16/05/2022	16/05/2022
							F1901	Additional Surgery	A070909	Oversensing	16/05/2022	16/05/2022
					CT2407	Active implantable	E2403	No Clinical Signs, Symptoms or Conditions	A070909	Oversensing	16/05/2022	16/05/2022
							F1901	Additional Surgery	A070909	Oversensing	16/05/2022	16/05/2022

					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A070909	Oversensing	16/05/2022	16/05/2022
							F1901	Additional Surgery	A070909	Oversensing	16/05/2022	16/05/2022
					CT2727	Active implant devices/ancillaries	E2403	No Clinical Signs, Symptoms or Conditions	A070909	Oversensing	16/05/2022	16/05/2022
							F1901	Additional Surgery	A070909	Oversensing	16/05/2022	16/05/2022
					CT2750	Implantable cardiac electrical stimulators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070909	Oversensing	16/05/2022	16/05/2022
							F1901	Additional Surgery	A070909	Oversensing	16/05/2022	16/05/2022
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A070909	Oversensing	16/05/2022	16/05/2022
							F1901	Additional Surgery	A070909	Oversensing	16/05/2022	16/05/2022
					CT2985	Active	E2403	No Clinical Signs, Symptoms or Conditions	A070909	Oversensing	16/05/2022	16/05/2022
							F1901	Additional Surgery	A070909	Oversensing	16/05/2022	16/05/2022
					CT3001	Implantable metal	E2403	No Clinical Signs, Symptoms or Conditions	A070909	Oversensing	16/05/2022	16/05/2022
							F1901	Additional Surgery	A070909	Oversensing	16/05/2022	16/05/2022
					CT3002	Implantable synthetic polymer	E2403	No Clinical Signs, Symptoms or Conditions	A070909	Oversensing	16/05/2022	16/05/2022
							F1901	Additional Surgery	A070909	Oversensing	16/05/2022	16/05/2022
					CT3006	Implantable inorganic material	E2403	No Clinical Signs, Symptoms or Conditions	A070909	Oversensing	16/05/2022	16/05/2022
							F1901	Additional Surgery	A070909	Oversensing	16/05/2022	16/05/2022
					CT314	Electrically powered	E2403	No Clinical Signs, Symptoms or Conditions	A070909	Oversensing	16/05/2022	16/05/2022
							F1901	Additional Surgery	A070909	Oversensing	16/05/2022	16/05/2022
					CT321	Long-term surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A070909	Oversensing	16/05/2022	16/05/2022
							F1901	Additional Surgery	A070909	Oversensing	16/05/2022	16/05/2022
					CT449	Generators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070909	Oversensing	16/05/2022	16/05/2022
							F1901	Additional Surgery	A070909	Oversensing	16/05/2022	16/05/2022
					CT981	Single use	E2403	No Clinical Signs, Symptoms or Conditions	A070909	Oversensing	16/05/2022	16/05/2022
							F1901	Additional Surgery	A070909	Oversensing	16/05/2022	16/05/2022
					CT983	Surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A070909	Oversensing	16/05/2022	16/05/2022
							F1901	Additional Surgery	A070909	Oversensing	16/05/2022	16/05/2022
6728	2022	All other reportable incidents	35223	Endocardial/interventricular septal pacing lead	CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A0401	Break		
									A070909	Oversensing		
							F26	No Health Consequences or Impact	A0401	Break		
									A070909	Oversensing		
6729	2021	All other reportable incidents	60910	Coronary venous pacing lead	CT1684	Cardiac pacemakers and associated devices	E012207	Twitching	A070102	High Capture Threshold	29/10/2021	29/10/2021
									A24	No Apparent Device Problem	29/10/2021	29/10/2021
									A27	Appropriate Device Problem Term/Code not Available	29/10/2021	29/10/2021
							E2403	No Clinical Signs, Symptoms or Conditions	A070102	High Capture Threshold	29/10/2021	29/10/2021
									A24	No Apparent Device Problem	29/10/2021	29/10/2021
									A27	Appropriate Device Problem Term/Code not Available	29/10/2021	29/10/2021
							F26	No Health Consequences or Impact	A070102	High Capture Threshold	29/10/2021	29/10/2021
									A24	No Apparent Device Problem	29/10/2021	29/10/2021
									A27	Appropriate Device Problem Term/Code not Available	29/10/2021	29/10/2021
6730	2021	All other reportable incidents	47270	Cardiac resynchronization therapy implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
6731	2021	All other reportable incidents	35852	Single-chamber implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem	10/11/2021	10/11/2021
									A070504	Premature Discharge of Battery	10/11/2021	10/11/2021
							F08	Hospitalization or Prolonged Hospitalization	A0705	Battery Problem	10/11/2021	10/11/2021
									A070504	Premature Discharge of Battery	10/11/2021	10/11/2021
							F1905	Device Revision or Replacement	A0705	Battery Problem	10/11/2021	10/11/2021
									A070504	Premature Discharge of Battery	10/11/2021	10/11/2021
6732	2021	All other reportable incidents	47265	Dual-chamber implantable pacemaker, rate-responsive	CT141	Electrodes/Leads and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070909	Oversensing		
							F26	No Health Consequences or Impact	A070909	Oversensing		
					CT145	Cardiovascular devices	E2403	No Clinical Signs, Symptoms or Conditions	A070909	Oversensing		
							F26	No Health Consequences or Impact	A070909	Oversensing		
					CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070909	Oversensing		
							F26	No Health Consequences or Impact	A070909	Oversensing		
					CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A070909	Oversensing		
							F26	No Health Consequences or Impact	A070909	Oversensing		
					CT233	Surgical	E2403	No Clinical Signs, Symptoms or Conditions	A070909	Oversensing		
							F26	No Health Consequences or Impact	A070909	Oversensing		
					CT2406	Implantable	E2403	No Clinical Signs, Symptoms or Conditions	A070909	Oversensing		
							F26	No Health Consequences or Impact	A070909	Oversensing		
					CT2407	Active implantable	E2403	No Clinical Signs, Symptoms or Conditions	A070909	Oversensing		
							F26	No Health Consequences or Impact	A070909	Oversensing		
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A070909	Oversensing		
							F26	No Health Consequences or Impact	A070909	Oversensing		
					CT2727	Active implant devices/ancillaries	E2403	No Clinical Signs, Symptoms or Conditions	A070909	Oversensing		
							F26	No Health Consequences or Impact	A070909	Oversensing		
					CT2750	Implantable cardiac electrical stimulators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070909	Oversensing		
							F26	No Health Consequences or Impact	A070909	Oversensing		
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A070909	Oversensing		
							F26	No Health Consequences or Impact	A070909	Oversensing		
					CT2985	Active	E2403	No Clinical Signs, Symptoms or Conditions	A070909	Oversensing		
							F26	No Health Consequences or Impact	A070909	Oversensing		
					CT3001	Implantable metal	E2403	No Clinical Signs, Symptoms or Conditions	A070909	Oversensing		
							F26	No Health Consequences or Impact	A070909	Oversensing		
					CT3002	Implantable synthetic polymer	E2403	No Clinical Signs, Symptoms or Conditions	A070909	Oversensing		
							F26	No Health Consequences or Impact	A070909	Oversensing		
					CT3006	Implantable inorganic material	E2403	No Clinical Signs, Symptoms or Conditions	A070909	Oversensing		
							F26	No Health Consequences or Impact	A070909	Oversensing		
					CT314	Electrically powered	E2403	No Clinical Signs, Symptoms or Conditions	A070909	Oversensing		
							F26	No Health Consequences or Impact	A070909	Oversensing		
					CT321	Long-term surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A070909	Oversensing		
							F26	No Health Consequences or Impact	A070909	Oversensing		
					CT449	Generators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070909	Oversensing		
							F26	No Health Consequences or Impact	A070909	Oversensing		
					CT981	Single use	E2403	No Clinical Signs, Symptoms or Conditions	A070909	Oversensing		
							F26	No Health Consequences or Impact	A070909	Oversensing		

					CT983	Surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A070909	Oversensing		
							F26	No Health Consequences or Impact	A070909	Oversensing		
6733	2022	All other reportable incidents	35223	Endocardial/interventricular septal pacing lead	CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A040611	Stretched		
									A07	Electrical /Electronic Property Problem		
									A070104	Unstable Capture Threshold		
									A070909	Oversensing		
									A0722	Impedance Problem		
									A072201	High Impedance		
							F23	Unexpected Medical Intervention	A040611	Stretched		
									A07	Electrical /Electronic Property Problem		
									A070104	Unstable Capture Threshold		
									A070909	Oversensing		
									A0722	Impedance Problem		
									A072201	High Impedance		
							F26	No Health Consequences or Impact	A040611	Stretched		
									A07	Electrical /Electronic Property Problem		
									A070104	Unstable Capture Threshold		
									A070909	Oversensing		
									A0722	Impedance Problem		
									A072201	High Impedance		
6734	2021	All other reportable incidents	47205	Cardiac pulse generator programmer	CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A0701	Capturing Problem		
									A0902	Display or Visual Feedback Problem		
									A11	Computer Software Problem		
									A1104	Computer Operating System Problem		
							F26	No Health Consequences or Impact	A0701	Capturing Problem		
									A0902	Display or Visual Feedback Problem		
									A11	Computer Software Problem		
									A1104	Computer Operating System Problem		
6735	2021	All other reportable incidents	47263	Cardiac resynchronization therapy implantable pacemaker	CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A0908	Incorrect, Inadequate or Imprecise Result or Readings	24/09/2021	24/09/2021
							F08	Hospitalization or Prolonged Hospitalization	A0908	Incorrect, Inadequate or Imprecise Result or Readings	24/09/2021	24/09/2021
							F1903	Device Explantation	A0908	Incorrect, Inadequate or Imprecise Result or Readings	24/09/2021	24/09/2021
							F1905	Device Revision or Replacement	A0908	Incorrect, Inadequate or Imprecise Result or Readings	24/09/2021	24/09/2021
					CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A0908	Incorrect, Inadequate or Imprecise Result or Readings	24/09/2021	24/09/2021
							F08	Hospitalization or Prolonged Hospitalization	A0908	Incorrect, Inadequate or Imprecise Result or Readings	24/09/2021	24/09/2021
							F1903	Device Explantation	A0908	Incorrect, Inadequate or Imprecise Result or Readings	24/09/2021	24/09/2021
							F1905	Device Revision or Replacement	A0908	Incorrect, Inadequate or Imprecise Result or Readings	24/09/2021	24/09/2021
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A0908	Incorrect, Inadequate or Imprecise Result or Readings	24/09/2021	24/09/2021
							F08	Hospitalization or Prolonged Hospitalization	A0908	Incorrect, Inadequate or Imprecise Result or Readings	24/09/2021	24/09/2021
							F1903	Device Explantation	A0908	Incorrect, Inadequate or Imprecise Result or Readings	24/09/2021	24/09/2021
							F1905	Device Revision or Replacement	A0908	Incorrect, Inadequate or Imprecise Result or Readings	24/09/2021	24/09/2021
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A0908	Incorrect, Inadequate or Imprecise Result or Readings	24/09/2021	24/09/2021
							F08	Hospitalization or Prolonged Hospitalization	A0908	Incorrect, Inadequate or Imprecise Result or Readings	24/09/2021	24/09/2021
							F1903	Device Explantation	A0908	Incorrect, Inadequate or Imprecise Result or Readings	24/09/2021	24/09/2021
							F1905	Device Revision or Replacement	A0908	Incorrect, Inadequate or Imprecise Result or Readings	24/09/2021	24/09/2021
6736	2021	All other reportable incidents	47265	Dual-chamber implantable pacemaker, rate-responsive	CT141	Electrodes/Leads and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F11	Non-Serious Injury/ Illness/ Impairment	A070504	Premature Discharge of Battery		
					CT145	Cardiovascular devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F11	Non-Serious Injury/ Illness/ Impairment	A070504	Premature Discharge of Battery		
					CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F11	Non-Serious Injury/ Illness/ Impairment	A070504	Premature Discharge of Battery		
					CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F11	Non-Serious Injury/ Illness/ Impairment	A070504	Premature Discharge of Battery		
					CT233	Surgical	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F11	Non-Serious Injury/ Illness/ Impairment	A070504	Premature Discharge of Battery		
					CT2406	Implantable	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F11	Non-Serious Injury/ Illness/ Impairment	A070504	Premature Discharge of Battery		
					CT2407	Active implantable	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F11	Non-Serious Injury/ Illness/ Impairment	A070504	Premature Discharge of Battery		
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F11	Non-Serious Injury/ Illness/ Impairment	A070504	Premature Discharge of Battery		
					CT2727	Active implant devices/ancillaries	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F11	Non-Serious Injury/ Illness/ Impairment	A070504	Premature Discharge of Battery		
					CT2750	Implantable cardiac electrical stimulators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F11	Non-Serious Injury/ Illness/ Impairment	A070504	Premature Discharge of Battery		
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F11	Non-Serious Injury/ Illness/ Impairment	A070504	Premature Discharge of Battery		
					CT2985	Active	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F11	Non-Serious Injury/ Illness/ Impairment	A070504	Premature Discharge of Battery		
					CT3001	Implantable metal	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F11	Non-Serious Injury/ Illness/ Impairment	A070504	Premature Discharge of Battery		
					CT3002	Implantable synthetic polymer	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F11	Non-Serious Injury/ Illness/ Impairment	A070504	Premature Discharge of Battery		
					CT3006	Implantable inorganic material	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F11	Non-Serious Injury/ Illness/ Impairment	A070504	Premature Discharge of Battery		
					CT314	Electrically powered	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F11	Non-Serious Injury/ Illness/ Impairment	A070504	Premature Discharge of Battery		

					CT321	Long-term surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F11	Non-Serious Injury/ Illness/ Impairment	A070504	Premature Discharge of Battery		
					CT449	Generators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F11	Non-Serious Injury/ Illness/ Impairment	A070504	Premature Discharge of Battery		
					CT981	Single use	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F11	Non-Serious Injury/ Illness/ Impairment	A070504	Premature Discharge of Battery		
					CT983	Surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F11	Non-Serious Injury/ Illness/ Impairment	A070504	Premature Discharge of Battery		
6737	2021	All other reportable incidents	35852	Single-chamber implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E0605	Cardiac Tamponade	A24	No Apparent Device Problem		
							E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem		
							F12	Serious Injury/ Illness/ Impairment	A24	No Apparent Device Problem		
							F1203	Life Threatening Illness or Injury	A24	No Apparent Device Problem		
							F1908	Prolonged Surgery	A24	No Apparent Device Problem		
6738	2021	All other reportable incidents	47265	Dual-chamber implantable pacemaker, rate-responsive	CT141	Electrodes/Leads and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
					CT145	Cardiovascular devices	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
					CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
					CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
					CT233	Surgical	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
					CT2406	Implantable	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
					CT2407	Active implantable	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
					CT2727	Active implant devices/ancillaries	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
					CT2750	Implantable cardiac electrical stimulators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
					CT2985	Active	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
					CT3001	Implantable metal	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
					CT3002	Implantable synthetic polymer	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
					CT3006	Implantable inorganic material	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
					CT314	Electrically powered	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
					CT321	Long-term surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
					CT449	Generators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
					CT981	Single use	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A0705	Battery Problem		

								A070504	Premature Discharge of Battery			
				CT983	Surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem			
								A070504	Premature Discharge of Battery			
						F26	No Health Consequences or Impact	A0705	Battery Problem			
								A070504	Premature Discharge of Battery			
6739	2022	All other reportable incidents	35223	Endocardial/interventricular septal pacing lead	CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070102	High Capture Threshold		
									A070901	Decreased Sensitivity		
									A090801	Signal Artifact/Noise		
							F26	No Health Consequences or Impact	A070102	High Capture Threshold		
									A070901	Decreased Sensitivity		
									A090801	Signal Artifact/Noise		
6740	2022	All other reportable incidents	36241	Pacemaker/defibrillator lead	CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A072201	High Impedance		
									A072202	Low Impedance		
							F08	Hospitalization or Prolonged Hospitalization	A072201	High Impedance		
									A072202	Low Impedance		
							F1905	Device Revision or Replacement	A072201	High Impedance		
									A072202	Low Impedance		
6741	2022	All other reportable incidents	36241	Pacemaker/defibrillator lead	CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A072201	High Impedance		
							F26	No Health Consequences or Impact	A072201	High Impedance		
6742	2022	All other reportable incidents	17882	Physiologic-monitoring defibrillation system	CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A150105	Key or Button Unresponsive/not Working		
							F27	Problem Identified During Non-Clinical Procedure	A150105	Key or Button Unresponsive/not Working		
6743	2022	All other reportable incidents	47265	Dual-chamber implantable pacemaker, rate-responsive	CT141	Electrodes/Leads and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	27/06/2022	27/06/2022
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	27/06/2022	27/06/2022
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	27/06/2022	27/06/2022
					CT145	Cardiovascular devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	27/06/2022	27/06/2022
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	27/06/2022	27/06/2022
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	27/06/2022	27/06/2022
					CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	27/06/2022	27/06/2022
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	27/06/2022	27/06/2022
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	27/06/2022	27/06/2022
					CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	27/06/2022	27/06/2022
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	27/06/2022	27/06/2022
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	27/06/2022	27/06/2022
					CT233	Surgical	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	27/06/2022	27/06/2022
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	27/06/2022	27/06/2022
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	27/06/2022	27/06/2022
					CT2406	Implantable	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	27/06/2022	27/06/2022
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	27/06/2022	27/06/2022
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	27/06/2022	27/06/2022
					CT2407	Active implantable	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	27/06/2022	27/06/2022
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	27/06/2022	27/06/2022
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	27/06/2022	27/06/2022
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	27/06/2022	27/06/2022
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	27/06/2022	27/06/2022
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	27/06/2022	27/06/2022
					CT2727	Active implant devices/ancillaries	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	27/06/2022	27/06/2022
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	27/06/2022	27/06/2022
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	27/06/2022	27/06/2022
					CT2750	Implantable cardiac electrical stimulators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	27/06/2022	27/06/2022
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	27/06/2022	27/06/2022
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	27/06/2022	27/06/2022
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	27/06/2022	27/06/2022
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	27/06/2022	27/06/2022
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	27/06/2022	27/06/2022
					CT2985	Active	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	27/06/2022	27/06/2022
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	27/06/2022	27/06/2022
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	27/06/2022	27/06/2022
					CT3001	Implantable metal	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	27/06/2022	27/06/2022
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	27/06/2022	27/06/2022
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	27/06/2022	27/06/2022
					CT3002	Implantable synthetic polymer	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	27/06/2022	27/06/2022
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	27/06/2022	27/06/2022
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	27/06/2022	27/06/2022
					CT3006	Implantable inorganic material	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	27/06/2022	27/06/2022
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	27/06/2022	27/06/2022
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	27/06/2022	27/06/2022
					CT314	Electrically powered	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	27/06/2022	27/06/2022
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	27/06/2022	27/06/2022
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	27/06/2022	27/06/2022
					CT321	Long-term surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	27/06/2022	27/06/2022
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	27/06/2022	27/06/2022
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	27/06/2022	27/06/2022
					CT449	Generators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	27/06/2022	27/06/2022
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	27/06/2022	27/06/2022
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	27/06/2022	27/06/2022
					CT981	Single use	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	27/06/2022	27/06/2022
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	27/06/2022	27/06/2022
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	27/06/2022	27/06/2022
					CT983	Surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	27/06/2022	27/06/2022
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	27/06/2022	27/06/2022
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	27/06/2022	27/06/2022
6744	2022	All other reportable incidents	47265	Dual-chamber implantable pacemaker, rate-responsive	CT141	Electrodes/Leads and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT145	Cardiovascular devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		

							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT233	Surgical	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT2406	Implantable	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT2407	Active implantable	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT2727	Active implant devices/ancillaries	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT2750	Implantable cardiac electrical stimulators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT2985	Active	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT3001	Implantable metal	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT3002	Implantable synthetic polymer	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT3006	Implantable inorganic material	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT314	Electrically powered	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT321	Long-term surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT449	Generators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT981	Single use	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT983	Surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
6745	2022	All other reportable incidents	47265	Dual-chamber implantable pacemaker, rate-responsive	CT141	Electrodes/Leads and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT145	Cardiovascular devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT233	Surgical	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT2406	Implantable	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT2407	Active implantable	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT2727	Active implant devices/ancillaries	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT2750	Implantable cardiac electrical stimulators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT2985	Active	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT3001	Implantable metal	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT3002	Implantable synthetic polymer	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT3006	Implantable inorganic material	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT314	Electrically powered	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT321	Long-term surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT449	Generators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT981	Single use	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT983	Surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
6746	2021	All other reportable incidents	35223	Endocardial/interventricular septal pacing lead	CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A0401	Break		
									A070909	Oversensing		
									A072202	Low Impedance		
									A090801	Signal Artifact/Noise		
							F26	No Health Consequences or Impact	A0401	Break		
									A070909	Oversensing		
									A072202	Low Impedance		
									A090801	Signal Artifact/Noise		
6747	2021	All other reportable incidents	36241	Pacemaker/defibrillator lead	CT1684	Cardiac pacemakers and associated devices	E2104	Electric Shock	A072201	High Impedance		
							E2403	No Clinical Signs, Symptoms or Conditions	A072201	High Impedance		
							F17	Sedation	A072201	High Impedance		
							F22	Unexpected Diagnostic Intervention	A072201	High Impedance		
6748	2022	All other reportable incidents	35223	Endocardial/interventricular septal pacing lead	CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070102	High Capture Threshold		
									A0722	Impedance Problem		
							F26	No Health Consequences or Impact	A070102	High Capture Threshold		

									A0722	Impedance Problem			
6749	2022	All other reportable incidents	63188	Left-ventricular wireless pacing system sensor/transmitter	CT141	Electrodes/Leads and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A0104	Migration or Expulsion of Device			
					CT145	Cardiovascular devices	E2403	No Clinical Signs, Symptoms or Conditions	A0104	Migration or Expulsion of Device			
					CT1577	Wireless monitoring system transmitters	E2403	No Clinical Signs, Symptoms or Conditions	A0104	Migration or Expulsion of Device			
					CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A0104	Migration or Expulsion of Device			
					CT2317	Transmitters/Receivers/Transceivers	E2403	No Clinical Signs, Symptoms or Conditions	A0104	Migration or Expulsion of Device			
					CT233	Surgical	E2403	No Clinical Signs, Symptoms or Conditions	A0104	Migration or Expulsion of Device			
					CT2406	Implantable	E2403	No Clinical Signs, Symptoms or Conditions	A0104	Migration or Expulsion of Device			
					CT2407	Active implantable	E2403	No Clinical Signs, Symptoms or Conditions	A0104	Migration or Expulsion of Device			
					CT255	Transducers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A0104	Migration or Expulsion of Device			
					CT2726	Physiological monitoring/measuring devices	E2403	No Clinical Signs, Symptoms or Conditions	A0104	Migration or Expulsion of Device			
					CT2727	Active implant devices/ancillaries	E2403	No Clinical Signs, Symptoms or Conditions	A0104	Migration or Expulsion of Device			
					CT2750	Implantable cardiac electrical stimulators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A0104	Migration or Expulsion of Device			
					CT2885	Physiological monitoring/measuring system transmitters/receivers/transceivers	E2403	No Clinical Signs, Symptoms or Conditions	A0104	Migration or Expulsion of Device			
					CT2985	Active	E2403	No Clinical Signs, Symptoms or Conditions	A0104	Migration or Expulsion of Device			
					CT3001	Implantable metal	E2403	No Clinical Signs, Symptoms or Conditions	A0104	Migration or Expulsion of Device			
					CT3002	Implantable synthetic polymer	E2403	No Clinical Signs, Symptoms or Conditions	A0104	Migration or Expulsion of Device			
					CT3006	Implantable inorganic material	E2403	No Clinical Signs, Symptoms or Conditions	A0104	Migration or Expulsion of Device			
					CT314	Electrically powered	E2403	No Clinical Signs, Symptoms or Conditions	A0104	Migration or Expulsion of Device			
					CT321	Long-term surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A0104	Migration or Expulsion of Device			
					CT384	Sensors and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A0104	Migration or Expulsion of Device			
					CT449	Generators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A0104	Migration or Expulsion of Device			
					CT981	Single use	E2403	No Clinical Signs, Symptoms or Conditions	A0104	Migration or Expulsion of Device			
					CT983	Surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A0104	Migration or Expulsion of Device			
					CT994	Sensors	E2403	No Clinical Signs, Symptoms or Conditions	A0104	Migration or Expulsion of Device			
6750	2022	All other reportable incidents	60910	Coronary venous pacing lead	CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A072202	Low Impedance			
							F23	Unexpected Medical Intervention	A072202	Low Impedance			
6751	2022	All other reportable incidents	37265	Dual-chamber implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A070909	Oversensing			
									A071302	Inappropriate/Inadequate Shock/Stimulation			
							F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A070909	Oversensing			
									A071302	Inappropriate/Inadequate Shock/Stimulation			
6752	2022	All other reportable incidents	47265	Dual-chamber implantable pacemaker, rate-responsive	CT141	Electrodes/Leads and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery			
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery			
					CT145	Cardiovascular devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery			
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery			
					CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery			
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery			
					CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery			
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery			
					CT233	Surgical	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery			
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery			
					CT2406	Implantable	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery			
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery			
					CT2407	Active implantable	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery			
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery			
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery			
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery			
					CT2727	Active implant devices/ancillaries	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery			
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery			
					CT2750	Implantable cardiac electrical stimulators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery			
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery			
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery			
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery			
					CT2985	Active	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery			
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery			
					CT3001	Implantable metal	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery			
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery			
					CT3002	Implantable synthetic polymer	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery			
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery			
					CT3006	Implantable inorganic material	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery			
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery			
					CT314	Electrically powered	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery			
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery			
					CT321	Long-term surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery			
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery			
					CT449	Generators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery			
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery			
					CT981	Single use	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery			
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery			
					CT983	Surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery			
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery			
6753	2021	All other reportable incidents	47265	Dual-chamber implantable pacemaker, rate-responsive	CT141	Electrodes/Leads and associated devices	E0112	Dizziness	A070101	Failure to Capture			
									A070909	Oversensing			
							F23	Unexpected Medical Intervention	A070101	Failure to Capture			
									A070909	Oversensing			
					CT145	Cardiovascular devices	E0112	Dizziness	A070101	Failure to Capture			
									A070909	Oversensing			
							F23	Unexpected Medical Intervention	A070101	Failure to Capture			
									A070909	Oversensing			
					CT1684	Cardiac pacemakers and associated devices	E0112	Dizziness	A070101	Failure to Capture			
									A070909	Oversensing			
							F23	Unexpected Medical Intervention	A070101	Failure to Capture			
									A070909	Oversensing			
					CT1685	Cardiac pacemakers	E0112	Dizziness	A070101	Failure to Capture			
									A070909	Oversensing			
							F23	Unexpected Medical Intervention	A070101	Failure to Capture			

								A070909	Oversensing			
				CT233	Surgical	E0112	Dizziness	A070101	Failure to Capture			
								A070909	Oversensing			
						F23	Unexpected Medical Intervention	A070101	Failure to Capture			
								A070909	Oversensing			
				CT2406	Implantable	E0112	Dizziness	A070101	Failure to Capture			
								A070909	Oversensing			
						F23	Unexpected Medical Intervention	A070101	Failure to Capture			
								A070909	Oversensing			
				CT2407	Active implantable	E0112	Dizziness	A070101	Failure to Capture			
								A070909	Oversensing			
						F23	Unexpected Medical Intervention	A070101	Failure to Capture			
								A070909	Oversensing			
				CT257	Implantable cardiac pacemakers	E0112	Dizziness	A070101	Failure to Capture			
								A070909	Oversensing			
						F23	Unexpected Medical Intervention	A070101	Failure to Capture			
								A070909	Oversensing			
				CT2727	Active implant devices/ancillaries	E0112	Dizziness	A070101	Failure to Capture			
								A070909	Oversensing			
						F23	Unexpected Medical Intervention	A070101	Failure to Capture			
								A070909	Oversensing			
				CT2750	Implantable cardiac electrical stimulators and associated devices	E0112	Dizziness	A070101	Failure to Capture			
								A070909	Oversensing			
						F23	Unexpected Medical Intervention	A070101	Failure to Capture			
								A070909	Oversensing			
				CT2866	Implantable cardiac electrical stimulators	E0112	Dizziness	A070101	Failure to Capture			
								A070909	Oversensing			
						F23	Unexpected Medical Intervention	A070101	Failure to Capture			
								A070909	Oversensing			
				CT2985	Active	E0112	Dizziness	A070101	Failure to Capture			
								A070909	Oversensing			
						F23	Unexpected Medical Intervention	A070101	Failure to Capture			
								A070909	Oversensing			
				CT3001	Implantable metal	E0112	Dizziness	A070101	Failure to Capture			
								A070909	Oversensing			
						F23	Unexpected Medical Intervention	A070101	Failure to Capture			
								A070909	Oversensing			
				CT3002	Implantable synthetic polymer	E0112	Dizziness	A070101	Failure to Capture			
								A070909	Oversensing			
						F23	Unexpected Medical Intervention	A070101	Failure to Capture			
								A070909	Oversensing			
				CT3006	Implantable inorganic material	E0112	Dizziness	A070101	Failure to Capture			
								A070909	Oversensing			
						F23	Unexpected Medical Intervention	A070101	Failure to Capture			
								A070909	Oversensing			
				CT314	Electrically powered	E0112	Dizziness	A070101	Failure to Capture			
								A070909	Oversensing			
						F23	Unexpected Medical Intervention	A070101	Failure to Capture			
								A070909	Oversensing			
				CT321	Long-term surgical invasive	E0112	Dizziness	A070101	Failure to Capture			
								A070909	Oversensing			
						F23	Unexpected Medical Intervention	A070101	Failure to Capture			
								A070909	Oversensing			
				CT449	Generators and associated devices	E0112	Dizziness	A070101	Failure to Capture			
								A070909	Oversensing			
						F23	Unexpected Medical Intervention	A070101	Failure to Capture			
								A070909	Oversensing			
				CT981	Single use	E0112	Dizziness	A070101	Failure to Capture			
								A070909	Oversensing			
						F23	Unexpected Medical Intervention	A070101	Failure to Capture			
								A070909	Oversensing			
				CT983	Surgical invasive	E0112	Dizziness	A070101	Failure to Capture			
								A070909	Oversensing			
						F23	Unexpected Medical Intervention	A070101	Failure to Capture			
								A070909	Oversensing			
6754	2021	All other reportable incidents	35852	Single-chamber implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E2104	Electric Shock	A070909	Oversensing		
									A071302	Inappropriate/Inadequate Shock/Stimulation		
									A1301	Failure to Read Input Signal		
						E2403	No Clinical Signs, Symptoms or Conditions	A070909	Oversensing			
								A071302	Inappropriate/Inadequate Shock/Stimulation			
								A1301	Failure to Read Input Signal			
						F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A070909	Oversensing			
								A071302	Inappropriate/Inadequate Shock/Stimulation			
								A1301	Failure to Read Input Signal			
6755	2022	All other reportable incidents	47265	Dual-chamber implantable pacemaker, rate-responsive	CT141	Electrodes/Leads and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070909	Oversensing	20/05/2022	20/05/2022
									A0712	Pacing Problem	20/05/2022	20/05/2022
						F1901	Additional Surgery	A070909	Oversensing	20/05/2022	20/05/2022	
								A0712	Pacing Problem	20/05/2022	20/05/2022	
				CT145	Cardiovascular devices	E2403	No Clinical Signs, Symptoms or Conditions	A070909	Oversensing	20/05/2022	20/05/2022	
								A0712	Pacing Problem	20/05/2022	20/05/2022	
						F1901	Additional Surgery	A070909	Oversensing	20/05/2022	20/05/2022	
								A0712	Pacing Problem	20/05/2022	20/05/2022	
				CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070909	Oversensing	20/05/2022	20/05/2022	
								A0712	Pacing Problem	20/05/2022	20/05/2022	
						F1901	Additional Surgery	A070909	Oversensing	20/05/2022	20/05/2022	
								A0712	Pacing Problem	20/05/2022	20/05/2022	
				CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A070909	Oversensing	20/05/2022	20/05/2022	
								A0712	Pacing Problem	20/05/2022	20/05/2022	

							F1901	Additional Surgery	A070909	Oversensing	20/05/2022	20/05/2022
									A0712	Pacing Problem	20/05/2022	20/05/2022
					CT233	Surgical	E2403	No Clinical Signs, Symptoms or Conditions	A070909	Oversensing	20/05/2022	20/05/2022
									A0712	Pacing Problem	20/05/2022	20/05/2022
							F1901	Additional Surgery	A070909	Oversensing	20/05/2022	20/05/2022
									A0712	Pacing Problem	20/05/2022	20/05/2022
					CT2406	Implantable	E2403	No Clinical Signs, Symptoms or Conditions	A070909	Oversensing	20/05/2022	20/05/2022
									A0712	Pacing Problem	20/05/2022	20/05/2022
							F1901	Additional Surgery	A070909	Oversensing	20/05/2022	20/05/2022
									A0712	Pacing Problem	20/05/2022	20/05/2022
					CT2407	Active implantable	E2403	No Clinical Signs, Symptoms or Conditions	A070909	Oversensing	20/05/2022	20/05/2022
									A0712	Pacing Problem	20/05/2022	20/05/2022
							F1901	Additional Surgery	A070909	Oversensing	20/05/2022	20/05/2022
									A0712	Pacing Problem	20/05/2022	20/05/2022
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A070909	Oversensing	20/05/2022	20/05/2022
									A0712	Pacing Problem	20/05/2022	20/05/2022
							F1901	Additional Surgery	A070909	Oversensing	20/05/2022	20/05/2022
									A0712	Pacing Problem	20/05/2022	20/05/2022
					CT2727	Active implant devices/ancillaries	E2403	No Clinical Signs, Symptoms or Conditions	A070909	Oversensing	20/05/2022	20/05/2022
									A0712	Pacing Problem	20/05/2022	20/05/2022
							F1901	Additional Surgery	A070909	Oversensing	20/05/2022	20/05/2022
									A0712	Pacing Problem	20/05/2022	20/05/2022
					CT2750	Implantable cardiac electrical stimulators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070909	Oversensing	20/05/2022	20/05/2022
									A0712	Pacing Problem	20/05/2022	20/05/2022
							F1901	Additional Surgery	A070909	Oversensing	20/05/2022	20/05/2022
									A0712	Pacing Problem	20/05/2022	20/05/2022
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A070909	Oversensing	20/05/2022	20/05/2022
									A0712	Pacing Problem	20/05/2022	20/05/2022
							F1901	Additional Surgery	A070909	Oversensing	20/05/2022	20/05/2022
									A0712	Pacing Problem	20/05/2022	20/05/2022
					CT2985	Active	E2403	No Clinical Signs, Symptoms or Conditions	A070909	Oversensing	20/05/2022	20/05/2022
									A0712	Pacing Problem	20/05/2022	20/05/2022
							F1901	Additional Surgery	A070909	Oversensing	20/05/2022	20/05/2022
									A0712	Pacing Problem	20/05/2022	20/05/2022
					CT3001	Implantable metal	E2403	No Clinical Signs, Symptoms or Conditions	A070909	Oversensing	20/05/2022	20/05/2022
									A0712	Pacing Problem	20/05/2022	20/05/2022
							F1901	Additional Surgery	A070909	Oversensing	20/05/2022	20/05/2022
									A0712	Pacing Problem	20/05/2022	20/05/2022
					CT3002	Implantable synthetic polymer	E2403	No Clinical Signs, Symptoms or Conditions	A070909	Oversensing	20/05/2022	20/05/2022
									A0712	Pacing Problem	20/05/2022	20/05/2022
							F1901	Additional Surgery	A070909	Oversensing	20/05/2022	20/05/2022
									A0712	Pacing Problem	20/05/2022	20/05/2022
					CT3006	Implantable inorganic material	E2403	No Clinical Signs, Symptoms or Conditions	A070909	Oversensing	20/05/2022	20/05/2022
									A0712	Pacing Problem	20/05/2022	20/05/2022
							F1901	Additional Surgery	A070909	Oversensing	20/05/2022	20/05/2022
									A0712	Pacing Problem	20/05/2022	20/05/2022
					CT314	Electrically powered	E2403	No Clinical Signs, Symptoms or Conditions	A070909	Oversensing	20/05/2022	20/05/2022
									A0712	Pacing Problem	20/05/2022	20/05/2022
							F1901	Additional Surgery	A070909	Oversensing	20/05/2022	20/05/2022
									A0712	Pacing Problem	20/05/2022	20/05/2022
					CT321	Long-term surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A070909	Oversensing	20/05/2022	20/05/2022
									A0712	Pacing Problem	20/05/2022	20/05/2022
							F1901	Additional Surgery	A070909	Oversensing	20/05/2022	20/05/2022
									A0712	Pacing Problem	20/05/2022	20/05/2022
					CT449	Generators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070909	Oversensing	20/05/2022	20/05/2022
									A0712	Pacing Problem	20/05/2022	20/05/2022
							F1901	Additional Surgery	A070909	Oversensing	20/05/2022	20/05/2022
									A0712	Pacing Problem	20/05/2022	20/05/2022
					CT981	Single use	E2403	No Clinical Signs, Symptoms or Conditions	A070909	Oversensing	20/05/2022	20/05/2022
									A0712	Pacing Problem	20/05/2022	20/05/2022
							F1901	Additional Surgery	A070909	Oversensing	20/05/2022	20/05/2022
									A0712	Pacing Problem	20/05/2022	20/05/2022
					CT983	Surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A070909	Oversensing	20/05/2022	20/05/2022
									A0712	Pacing Problem	20/05/2022	20/05/2022
							F1901	Additional Surgery	A070909	Oversensing	20/05/2022	20/05/2022
									A0712	Pacing Problem	20/05/2022	20/05/2022
6756	2022	All other reportable incidents	47270	Cardiac resynchronization therapy implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	23/05/2022	23/05/2022
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	23/05/2022	23/05/2022
							F12	Serious Injury/ Illness/ Impairment	A070504	Premature Discharge of Battery	23/05/2022	23/05/2022
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	23/05/2022	23/05/2022
6757	2022	All other reportable incidents	37265	Dual-chamber implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A0203	Defective Device		
							F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A0203	Defective Device		
6758	2022	All other reportable incidents	47267	Single-chamber implantable pacemaker, rate-responsive	CT141	Electrodes/Leads and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070101	Failure to Capture		
							F26	No Health Consequences or Impact	A070101	Failure to Capture		
					CT145	Cardiovascular devices	E2403	No Clinical Signs, Symptoms or Conditions	A070101	Failure to Capture		
							F26	No Health Consequences or Impact	A070101	Failure to Capture		
					CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070101	Failure to Capture		
							F26	No Health Consequences or Impact	A070101	Failure to Capture		
					CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A070101	Failure to Capture		
							F26	No Health Consequences or Impact	A070101	Failure to Capture		
					CT233	Surgical	E2403	No Clinical Signs, Symptoms or Conditions	A070101	Failure to Capture		
							F26	No Health Consequences or Impact	A070101	Failure to Capture		
					CT2406	Implantable	E2403	No Clinical Signs, Symptoms or Conditions	A070101	Failure to Capture		
							F26	No Health Consequences or Impact	A070101	Failure to Capture		
					CT2407	Active implantable	E2403	No Clinical Signs, Symptoms or Conditions	A070101	Failure to Capture		
							F26	No Health Consequences or Impact	A070101	Failure to Capture		
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A070101	Failure to Capture		
							F26	No Health Consequences or Impact	A070101	Failure to Capture		

					CT2727	Active implant devices/ancillaries	E2403	No Clinical Signs, Symptoms or Conditions	A070101	Failure to Capture		
							F26	No Health Consequences or Impact	A070101	Failure to Capture		
					CT2750	Implantable cardiac electrical stimulators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070101	Failure to Capture		
							F26	No Health Consequences or Impact	A070101	Failure to Capture		
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A070101	Failure to Capture		
							F26	No Health Consequences or Impact	A070101	Failure to Capture		
					CT2985	Active	E2403	No Clinical Signs, Symptoms or Conditions	A070101	Failure to Capture		
							F26	No Health Consequences or Impact	A070101	Failure to Capture		
					CT3001	Implantable metal	E2403	No Clinical Signs, Symptoms or Conditions	A070101	Failure to Capture		
							F26	No Health Consequences or Impact	A070101	Failure to Capture		
					CT3002	Implantable synthetic polymer	E2403	No Clinical Signs, Symptoms or Conditions	A070101	Failure to Capture		
							F26	No Health Consequences or Impact	A070101	Failure to Capture		
					CT3006	Implantable inorganic material	E2403	No Clinical Signs, Symptoms or Conditions	A070101	Failure to Capture		
							F26	No Health Consequences or Impact	A070101	Failure to Capture		
					CT314	Electrically powered	E2403	No Clinical Signs, Symptoms or Conditions	A070101	Failure to Capture		
							F26	No Health Consequences or Impact	A070101	Failure to Capture		
					CT321	Long-term surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A070101	Failure to Capture		
							F26	No Health Consequences or Impact	A070101	Failure to Capture		
					CT449	Generators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070101	Failure to Capture		
							F26	No Health Consequences or Impact	A070101	Failure to Capture		
					CT981	Single use	E2403	No Clinical Signs, Symptoms or Conditions	A070101	Failure to Capture		
							F26	No Health Consequences or Impact	A070101	Failure to Capture		
					CT983	Surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A070101	Failure to Capture		
							F26	No Health Consequences or Impact	A070101	Failure to Capture		
6759	2022	All other reportable incidents	35852	Single-chamber implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem	21/06/2022	22/06/2022
									A070504	Premature Discharge of Battery	21/06/2022	22/06/2022
							F08	Hospitalization or Prolonged Hospitalization	A0705	Battery Problem	21/06/2022	22/06/2022
									A070504	Premature Discharge of Battery	21/06/2022	22/06/2022
							F1905	Device Revision or Replacement	A0705	Battery Problem	21/06/2022	22/06/2022
									A070504	Premature Discharge of Battery	21/06/2022	22/06/2022
6760	2022	All other reportable incidents	47205	Cardiac pulse generator programmer	CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A0203	Defective Device		
									A150105	Key or Button Unresponsive/not Working		
							F1908	Prolonged Surgery	A0203	Defective Device		
									A150105	Key or Button Unresponsive/not Working		
6761	2021	All other reportable incidents	35223	Endocardial/interventricular septal pacing lead	CT1684	Cardiac pacemakers and associated devices	E2114	Perforation	A0701	Capturing Problem		
									A0722	Impedance Problem		
							E2403	No Clinical Signs, Symptoms or Conditions	A0701	Capturing Problem		
									A0722	Impedance Problem		
							F1203	Life Threatening Illness or Injury	A0701	Capturing Problem		
									A0722	Impedance Problem		
6762	2021	All other reportable incidents	35223	Endocardial/interventricular septal pacing lead	CT1684	Cardiac pacemakers and associated devices	E0602	Cardiac Arrest	A051201	Device Dislodged or Dislocated		
							E2403	No Clinical Signs, Symptoms or Conditions	A051201	Device Dislodged or Dislocated		
							F08	Hospitalization or Prolonged Hospitalization	A051201	Device Dislodged or Dislocated		
							F1203	Life Threatening Illness or Injury	A051201	Device Dislodged or Dislocated		
							F1904	Device Repositioning	A051201	Device Dislodged or Dislocated		
6763	2021	All other reportable incidents	17882	Physiologic-monitoring defibrillation system	CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070803	Failure to Power Up		
							F27	Problem Identified During Non-Clinical Procedure	A070803	Failure to Power Up		
6764	2021	All other reportable incidents	17882	Physiologic-monitoring defibrillation system	CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070803	Failure to Power Up		
							F27	Problem Identified During Non-Clinical Procedure	A070803	Failure to Power Up		
6765	2021	All other reportable incidents	17882	Physiologic-monitoring defibrillation system	CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070803	Failure to Power Up		
							F27	Problem Identified During Non-Clinical Procedure	A070803	Failure to Power Up		
6766	2021	All other reportable incidents	17882	Physiologic-monitoring defibrillation system	CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070803	Failure to Power Up		
							F27	Problem Identified During Non-Clinical Procedure	A070803	Failure to Power Up		
6767	2021	All other reportable incidents	17882	Physiologic-monitoring defibrillation system	CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070803	Failure to Power Up		
							F27	Problem Identified During Non-Clinical Procedure	A070803	Failure to Power Up		
6768	2021	All other reportable incidents	17882	Physiologic-monitoring defibrillation system	CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070803	Failure to Power Up		
							F27	Problem Identified During Non-Clinical Procedure	A070803	Failure to Power Up		
6769	2021	All other reportable incidents	17882	Physiologic-monitoring defibrillation system	CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070803	Failure to Power Up		
							F27	Problem Identified During Non-Clinical Procedure	A070803	Failure to Power Up		
6770	2021	All other reportable incidents	17882	Physiologic-monitoring defibrillation system	CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070803	Failure to Power Up		
							F27	Problem Identified During Non-Clinical Procedure	A070803	Failure to Power Up		
6771	2021	All other reportable incidents	35852	Single-chamber implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A0508	Noise, Audible		
									A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A0508	Noise, Audible		
									A070504	Premature Discharge of Battery		
6772	2021	All other reportable incidents	36241	Pacemaker/defibrillator lead	CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A072201	High Impedance		
							F2203	Imaging Required	A072201	High Impedance		
							F26	No Health Consequences or Impact	A072201	High Impedance		
6773	2021	All other reportable incidents	35852	Single-chamber implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem	08/11/2021	08/11/2021
									A070504	Premature Discharge of Battery	08/11/2021	08/11/2021
							F08	Hospitalization or Prolonged Hospitalization	A0705	Battery Problem	08/11/2021	08/11/2021
									A070504	Premature Discharge of Battery	08/11/2021	08/11/2021
							F1905	Device Revision or Replacement	A0705	Battery Problem	08/11/2021	08/11/2021
									A070504	Premature Discharge of Battery	08/11/2021	08/11/2021
6774	2021	All other reportable incidents	47265	Dual-chamber implantable pacemaker, rate-responsive	CT141	Electrodes/Leads and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	08/02/2021	08/02/2021
							F19	Surgical Intervention	A24	No Apparent Device Problem	08/02/2021	08/02/2021
					CT145	Cardiovascular devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	08/02/2021	08/02/2021
							F19	Surgical Intervention	A24	No Apparent Device Problem	08/02/2021	08/02/2021
					CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	08/02/2021	08/02/2021
							F19	Surgical Intervention	A24	No Apparent Device Problem	08/02/2021	08/02/2021
					CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	08/02/2021	08/02/2021
							F19	Surgical Intervention	A24	No Apparent Device Problem	08/02/2021	08/02/2021
					CT233	Surgical	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	08/02/2021	08/02/2021
							F19	Surgical Intervention	A24	No Apparent Device Problem	08/02/2021	08/02/2021
					CT2406	Implantable	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	08/02/2021	08/02/2021
							F19	Surgical Intervention	A24	No Apparent Device Problem	08/02/2021	08/02/2021
					CT2407	Active implantable	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	08/02/2021	08/02/2021
							F19	Surgical Intervention	A24	No Apparent Device Problem	08/02/2021	08/02/2021

					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	08/02/2021	08/02/2021
							F19	Surgical Intervention	A24	No Apparent Device Problem	08/02/2021	08/02/2021
					CT2727	Active implant devices/ancillaries	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	08/02/2021	08/02/2021
							F19	Surgical Intervention	A24	No Apparent Device Problem	08/02/2021	08/02/2021
					CT2750	Implantable cardiac electrical stimulators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	08/02/2021	08/02/2021
							F19	Surgical Intervention	A24	No Apparent Device Problem	08/02/2021	08/02/2021
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	08/02/2021	08/02/2021
							F19	Surgical Intervention	A24	No Apparent Device Problem	08/02/2021	08/02/2021
					CT2985	Active	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	08/02/2021	08/02/2021
							F19	Surgical Intervention	A24	No Apparent Device Problem	08/02/2021	08/02/2021
					CT3001	Implantable metal	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	08/02/2021	08/02/2021
							F19	Surgical Intervention	A24	No Apparent Device Problem	08/02/2021	08/02/2021
					CT3002	Implantable synthetic polymer	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	08/02/2021	08/02/2021
							F19	Surgical Intervention	A24	No Apparent Device Problem	08/02/2021	08/02/2021
					CT3006	Implantable inorganic material	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	08/02/2021	08/02/2021
							F19	Surgical Intervention	A24	No Apparent Device Problem	08/02/2021	08/02/2021
					CT314	Electrically powered	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	08/02/2021	08/02/2021
							F19	Surgical Intervention	A24	No Apparent Device Problem	08/02/2021	08/02/2021
					CT321	Long-term surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	08/02/2021	08/02/2021
							F19	Surgical Intervention	A24	No Apparent Device Problem	08/02/2021	08/02/2021
					CT449	Generators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	08/02/2021	08/02/2021
							F19	Surgical Intervention	A24	No Apparent Device Problem	08/02/2021	08/02/2021
					CT981	Single use	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	08/02/2021	08/02/2021
							F19	Surgical Intervention	A24	No Apparent Device Problem	08/02/2021	08/02/2021
					CT983	Surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	08/02/2021	08/02/2021
							F19	Surgical Intervention	A24	No Apparent Device Problem	08/02/2021	08/02/2021
6775	2022	All other reportable incidents	37265	Dual-chamber implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
6776	2022	All other reportable incidents	35223	Endocardial/interventricular septal pacing lead	CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A07	Electrical /Electronic Property Problem		
									A072201	High Impedance		
							F12	Serious Injury/ Illness/ Impairment	A07	Electrical /Electronic Property Problem		
									A072201	High Impedance		
							F23	Unexpected Medical Intervention	A07	Electrical /Electronic Property Problem		
									A072201	High Impedance		
							F26	No Health Consequences or Impact	A07	Electrical /Electronic Property Problem		
									A072201	High Impedance		
6777	2022	All other reportable incidents	35852	Single-chamber implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem	24/05/2022	24/05/2022
									A070504	Premature Discharge of Battery	24/05/2022	24/05/2022
							F08	Hospitalization or Prolonged Hospitalization	A0705	Battery Problem	24/05/2022	24/05/2022
									A070504	Premature Discharge of Battery	24/05/2022	24/05/2022
							F1905	Device Revision or Replacement	A0705	Battery Problem	24/05/2022	24/05/2022
									A070504	Premature Discharge of Battery	24/05/2022	24/05/2022
6778	2022	All other reportable incidents	17882	Physiologic-monitoring defibrillation system	CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A07	Electrical /Electronic Property Problem		
							F27	Problem Identified During Non-Clinical Procedure	A07	Electrical /Electronic Property Problem		
6779	2021	All other reportable incidents	17882	Physiologic-monitoring defibrillation system	CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A0719	Unexpected Shutdown		
							F26	No Health Consequences or Impact	A0719	Unexpected Shutdown		
6780	2021	All other reportable incidents	17882	Physiologic-monitoring defibrillation system	CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A090811	Unable to Obtain Readings		
							F27	Problem Identified During Non-Clinical Procedure	A090811	Unable to Obtain Readings		
6781	2022	All other reportable incidents	17882	Physiologic-monitoring defibrillation system	CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A0710	Failure to Conduct		
							F27	Problem Identified During Non-Clinical Procedure	A0710	Failure to Conduct		
6782	2021	All other reportable incidents	35852	Single-chamber implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A0508	Noise, Audible		
									A0908	Incorrect, Inadequate or Imprecise Result or Readings		
							F26	No Health Consequences or Impact	A0508	Noise, Audible		
									A0908	Incorrect, Inadequate or Imprecise Result or Readings		
6783	2022	All other reportable incidents	35223	Endocardial/interventricular septal pacing lead	CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070909	Oversensing		
									A070910	Undersensing		
							F26	No Health Consequences or Impact	A070909	Oversensing		
									A070910	Undersensing		
6784	2022	All other reportable incidents	17882	Physiologic-monitoring defibrillation system	CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A07	Electrical /Electronic Property Problem		
							F27	Problem Identified During Non-Clinical Procedure	A07	Electrical /Electronic Property Problem		
6785	2022	All other reportable incidents	17882	Physiologic-monitoring defibrillation system	CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A071301	Failure to Deliver Shock/Stimulation		
							F26	No Health Consequences or Impact	A071301	Failure to Deliver Shock/Stimulation		
6786	2021	All other reportable incidents	47265	Dual-chamber implantable pacemaker, rate-responsive	CT141	Electrodes/Leads and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	06/10/2021	06/10/2021
							F1901	Additional Surgery	A070504	Premature Discharge of Battery	06/10/2021	06/10/2021
					CT145	Cardiovascular devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	06/10/2021	06/10/2021
							F1901	Additional Surgery	A070504	Premature Discharge of Battery	06/10/2021	06/10/2021
					CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	06/10/2021	06/10/2021
							F1901	Additional Surgery	A070504	Premature Discharge of Battery	06/10/2021	06/10/2021
					CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	06/10/2021	06/10/2021
							F1901	Additional Surgery	A070504	Premature Discharge of Battery	06/10/2021	06/10/2021
					CT233	Surgical	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	06/10/2021	06/10/2021
							F1901	Additional Surgery	A070504	Premature Discharge of Battery	06/10/2021	06/10/2021
					CT2406	Implantable	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	06/10/2021	06/10/2021
							F1901	Additional Surgery	A070504	Premature Discharge of Battery	06/10/2021	06/10/2021
					CT2407	Active implantable	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	06/10/2021	06/10/2021
							F1901	Additional Surgery	A070504	Premature Discharge of Battery	06/10/2021	06/10/2021
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	06/10/2021	06/10/2021
							F1901	Additional Surgery	A070504	Premature Discharge of Battery	06/10/2021	06/10/2021
					CT2727	Active implant devices/ancillaries	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	06/10/2021	06/10/2021
							F1901	Additional Surgery	A070504	Premature Discharge of Battery	06/10/2021	06/10/2021
					CT2750	Implantable cardiac electrical stimulators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	06/10/2021	06/10/2021
							F1901	Additional Surgery	A070504	Premature Discharge of Battery	06/10/2021	06/10/2021
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	06/10/2021	06/10/2021
							F1901	Additional Surgery	A070504	Premature Discharge of Battery	06/10/2021	06/10/2021

					CT2985	Active	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	06/10/2021	06/10/2021
							F1901	Additional Surgery	A070504	Premature Discharge of Battery	06/10/2021	06/10/2021
					CT3001	Implantable metal	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	06/10/2021	06/10/2021
							F1901	Additional Surgery	A070504	Premature Discharge of Battery	06/10/2021	06/10/2021
					CT3002	Implantable synthetic polymer	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	06/10/2021	06/10/2021
							F1901	Additional Surgery	A070504	Premature Discharge of Battery	06/10/2021	06/10/2021
					CT3006	Implantable inorganic material	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	06/10/2021	06/10/2021
							F1901	Additional Surgery	A070504	Premature Discharge of Battery	06/10/2021	06/10/2021
					CT314	Electrically powered	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	06/10/2021	06/10/2021
							F1901	Additional Surgery	A070504	Premature Discharge of Battery	06/10/2021	06/10/2021
					CT321	Long-term surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	06/10/2021	06/10/2021
							F1901	Additional Surgery	A070504	Premature Discharge of Battery	06/10/2021	06/10/2021
					CT449	Generators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	06/10/2021	06/10/2021
							F1901	Additional Surgery	A070504	Premature Discharge of Battery	06/10/2021	06/10/2021
					CT981	Single use	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	06/10/2021	06/10/2021
							F1901	Additional Surgery	A070504	Premature Discharge of Battery	06/10/2021	06/10/2021
					CT983	Surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	06/10/2021	06/10/2021
							F1901	Additional Surgery	A070504	Premature Discharge of Battery	06/10/2021	06/10/2021
6787	2021	All other reportable incidents	35852	Single-chamber implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem	18/11/2021	18/11/2021
									A070504	Premature Discharge of Battery	18/11/2021	18/11/2021
							F08	Hospitalization or Prolonged Hospitalization	A0705	Battery Problem	18/11/2021	18/11/2021
									A070504	Premature Discharge of Battery	18/11/2021	18/11/2021
							F1905	Device Revision or Replacement	A0705	Battery Problem	18/11/2021	18/11/2021
									A070504	Premature Discharge of Battery	18/11/2021	18/11/2021
6788	2022	All other reportable incidents	35852	Single-chamber implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem	15/06/2022	15/06/2022
									A070504	Premature Discharge of Battery	15/06/2022	15/06/2022
							F08	Hospitalization or Prolonged Hospitalization	A0705	Battery Problem	15/06/2022	15/06/2022
									A070504	Premature Discharge of Battery	15/06/2022	15/06/2022
							F1903	Device Explantation	A0705	Battery Problem	15/06/2022	15/06/2022
									A070504	Premature Discharge of Battery	15/06/2022	15/06/2022
6789	2021	All other reportable incidents	35223	Endocardial/interventricular septal pacing lead	CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070909	Oversensing		
									A090801	Signal Artifact/Noise		
							F26	No Health Consequences or Impact	A070909	Oversensing		
									A090801	Signal Artifact/Noise		
6790	2022	All other reportable incidents	35852	Single-chamber implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem	22/07/2022	22/07/2022
									A070504	Premature Discharge of Battery	22/07/2022	22/07/2022
							F08	Hospitalization or Prolonged Hospitalization	A0705	Battery Problem	22/07/2022	22/07/2022
									A070504	Premature Discharge of Battery	22/07/2022	22/07/2022
							F1905	Device Revision or Replacement	A0705	Battery Problem	22/07/2022	22/07/2022
									A070504	Premature Discharge of Battery	22/07/2022	22/07/2022
6791	2021	All other reportable incidents	47265	Dual-chamber implantable pacemaker, rate-responsive	CT141	Electrodes/Leads and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	04/10/2021	04/10/2021
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	04/10/2021	04/10/2021
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	04/10/2021	04/10/2021
					CT145	Cardiovascular devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	04/10/2021	04/10/2021
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	04/10/2021	04/10/2021
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	04/10/2021	04/10/2021
					CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	04/10/2021	04/10/2021
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	04/10/2021	04/10/2021
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	04/10/2021	04/10/2021
					CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	04/10/2021	04/10/2021
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	04/10/2021	04/10/2021
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	04/10/2021	04/10/2021
					CT233	Surgical	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	04/10/2021	04/10/2021
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	04/10/2021	04/10/2021
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	04/10/2021	04/10/2021
					CT2406	Implantable	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	04/10/2021	04/10/2021
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	04/10/2021	04/10/2021
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	04/10/2021	04/10/2021
					CT2407	Active implantable	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	04/10/2021	04/10/2021
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	04/10/2021	04/10/2021
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	04/10/2021	04/10/2021
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	04/10/2021	04/10/2021
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	04/10/2021	04/10/2021
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	04/10/2021	04/10/2021
					CT2727	Active implant devices/ancillaries	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	04/10/2021	04/10/2021
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	04/10/2021	04/10/2021
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	04/10/2021	04/10/2021
					CT2750	Implantable cardiac electrical stimulators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	04/10/2021	04/10/2021
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	04/10/2021	04/10/2021
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	04/10/2021	04/10/2021
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	04/10/2021	04/10/2021
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	04/10/2021	04/10/2021
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	04/10/2021	04/10/2021
					CT2985	Active	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	04/10/2021	04/10/2021
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	04/10/2021	04/10/2021
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	04/10/2021	04/10/2021
					CT3001	Implantable metal	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	04/10/2021	04/10/2021
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	04/10/2021	04/10/2021
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	04/10/2021	04/10/2021
					CT3002	Implantable synthetic polymer	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	04/10/2021	04/10/2021
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	04/10/2021	04/10/2021
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	04/10/2021	04/10/2021
					CT3006	Implantable inorganic material	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	04/10/2021	04/10/2021
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	04/10/2021	04/10/2021
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	04/10/2021	04/10/2021
					CT314	Electrically powered	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	04/10/2021	04/10/2021
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	04/10/2021	04/10/2021
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	04/10/2021	04/10/2021

					CT321	Long-term surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	04/10/2021	04/10/2021
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	04/10/2021	04/10/2021
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	04/10/2021	04/10/2021
					CT449	Generators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	04/10/2021	04/10/2021
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	04/10/2021	04/10/2021
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	04/10/2021	04/10/2021
					CT981	Single use	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	04/10/2021	04/10/2021
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	04/10/2021	04/10/2021
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	04/10/2021	04/10/2021
					CT983	Surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	04/10/2021	04/10/2021
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	04/10/2021	04/10/2021
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	04/10/2021	04/10/2021
6792	2021	All other reportable incidents	36241	Pacemaker/defibrillator lead	CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A072201	High Impedance		
							F26	No Health Consequences or Impact	A072201	High Impedance		
6793	2022	All other reportable incidents	37265	Dual-chamber implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A070909	Oversensing		
							F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A070909	Oversensing		
6794	2022	All other reportable incidents	47265	Dual-chamber implantable pacemaker, rate-responsive	CT141	Electrodes/Leads and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
					CT145	Cardiovascular devices	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
					CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
					CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
					CT233	Surgical	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
					CT2406	Implantable	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
					CT2407	Active implantable	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
					CT2727	Active implant devices/ancillaries	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
					CT2750	Implantable cardiac electrical stimulators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
					CT2985	Active	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
					CT3001	Implantable metal	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
					CT3002	Implantable synthetic polymer	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
					CT3006	Implantable inorganic material	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
					CT314	Electrically powered	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
					CT321	Long-term surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
					CT449	Generators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		

					CT981	Single use	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
					CT983	Surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
6795	2022	All other reportable incidents	47265	Dual-chamber implantable pacemaker, rate-responsive	CT141	Electrodes/Leads and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A0203	Defective Device		
									A070101	Failure to Capture		
							F26	No Health Consequences or Impact	A0203	Defective Device		
									A070101	Failure to Capture		
					CT145	Cardiovascular devices	E2403	No Clinical Signs, Symptoms or Conditions	A0203	Defective Device		
									A070101	Failure to Capture		
							F26	No Health Consequences or Impact	A0203	Defective Device		
									A070101	Failure to Capture		
					CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A0203	Defective Device		
									A070101	Failure to Capture		
							F26	No Health Consequences or Impact	A0203	Defective Device		
									A070101	Failure to Capture		
					CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A0203	Defective Device		
									A070101	Failure to Capture		
							F26	No Health Consequences or Impact	A0203	Defective Device		
									A070101	Failure to Capture		
					CT233	Surgical	E2403	No Clinical Signs, Symptoms or Conditions	A0203	Defective Device		
									A070101	Failure to Capture		
							F26	No Health Consequences or Impact	A0203	Defective Device		
									A070101	Failure to Capture		
					CT2406	Implantable	E2403	No Clinical Signs, Symptoms or Conditions	A0203	Defective Device		
									A070101	Failure to Capture		
							F26	No Health Consequences or Impact	A0203	Defective Device		
									A070101	Failure to Capture		
					CT2407	Active implantable	E2403	No Clinical Signs, Symptoms or Conditions	A0203	Defective Device		
									A070101	Failure to Capture		
							F26	No Health Consequences or Impact	A0203	Defective Device		
									A070101	Failure to Capture		
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A0203	Defective Device		
									A070101	Failure to Capture		
							F26	No Health Consequences or Impact	A0203	Defective Device		
									A070101	Failure to Capture		
					CT2727	Active implant devices/ancillaries	E2403	No Clinical Signs, Symptoms or Conditions	A0203	Defective Device		
									A070101	Failure to Capture		
							F26	No Health Consequences or Impact	A0203	Defective Device		
									A070101	Failure to Capture		
					CT2750	Implantable cardiac electrical stimulators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A0203	Defective Device		
									A070101	Failure to Capture		
							F26	No Health Consequences or Impact	A0203	Defective Device		
									A070101	Failure to Capture		
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A0203	Defective Device		
									A070101	Failure to Capture		
							F26	No Health Consequences or Impact	A0203	Defective Device		
									A070101	Failure to Capture		
					CT2985	Active	E2403	No Clinical Signs, Symptoms or Conditions	A0203	Defective Device		
									A070101	Failure to Capture		
							F26	No Health Consequences or Impact	A0203	Defective Device		
									A070101	Failure to Capture		
					CT3001	Implantable metal	E2403	No Clinical Signs, Symptoms or Conditions	A0203	Defective Device		
									A070101	Failure to Capture		
							F26	No Health Consequences or Impact	A0203	Defective Device		
									A070101	Failure to Capture		
					CT3002	Implantable synthetic polymer	E2403	No Clinical Signs, Symptoms or Conditions	A0203	Defective Device		
									A070101	Failure to Capture		
							F26	No Health Consequences or Impact	A0203	Defective Device		
									A070101	Failure to Capture		
					CT3006	Implantable inorganic material	E2403	No Clinical Signs, Symptoms or Conditions	A0203	Defective Device		
									A070101	Failure to Capture		
							F26	No Health Consequences or Impact	A0203	Defective Device		
									A070101	Failure to Capture		
					CT314	Electrically powered	E2403	No Clinical Signs, Symptoms or Conditions	A0203	Defective Device		
									A070101	Failure to Capture		
							F26	No Health Consequences or Impact	A0203	Defective Device		
									A070101	Failure to Capture		
					CT321	Long-term surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A0203	Defective Device		
									A070101	Failure to Capture		
							F26	No Health Consequences or Impact	A0203	Defective Device		
									A070101	Failure to Capture		
					CT449	Generators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A0203	Defective Device		
									A070101	Failure to Capture		
							F26	No Health Consequences or Impact	A0203	Defective Device		
									A070101	Failure to Capture		
					CT981	Single use	E2403	No Clinical Signs, Symptoms or Conditions	A0203	Defective Device		
									A070101	Failure to Capture		
							F26	No Health Consequences or Impact	A0203	Defective Device		
									A070101	Failure to Capture		
					CT983	Surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A0203	Defective Device		
									A070101	Failure to Capture		
							F26	No Health Consequences or Impact	A0203	Defective Device		
									A070101	Failure to Capture		

6796	2022	All other reportable incidents	60910	Coronary venous pacing lead	CT1684	Cardiac pacemakers and associated devices	E0717	Dyspnea	A070102	High Capture Threshold		
									A0712	Pacing Problem		
									A072202	Low Impedance		
									A15	Activation, Positioning or Separation Problem		
									A2303	Improper or Incorrect Procedure or Method		
							E2403	No Clinical Signs, Symptoms or Conditions	A070102	High Capture Threshold		
									A0712	Pacing Problem		
									A072202	Low Impedance		
									A15	Activation, Positioning or Separation Problem		
									A2303	Improper or Incorrect Procedure or Method		
							F11	Non-Serious Injury/ Illness/ Impairment	A070102	High Capture Threshold		
									A0712	Pacing Problem		
									A072202	Low Impedance		
									A15	Activation, Positioning or Separation Problem		
									A2303	Improper or Incorrect Procedure or Method		
6797	2021	All other reportable incidents	35223	Endocardial/interventricular septal pacing lead	CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A0401	Break		
									A070101	Failure to Capture		
									A070909	Oversensing		
									A072202	Low Impedance		
							F1901	Additional Surgery	A0401	Break		
									A070101	Failure to Capture		
									A070909	Oversensing		
									A072202	Low Impedance		
6798	2021	All other reportable incidents	37265	Dual-chamber implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E0601	Arrhythmia	A070910	Undersensing		
							E0742	Respiratory Failure	A070910	Undersensing		
							E2402	Appropriate Clinical Signs, Symptoms and Conditions Term/Code not Available	A070910	Undersensing		
							E2403	No Clinical Signs, Symptoms or Conditions	A070910	Undersensing		
							F1203	Life Threatening Illness or Injury	A070910	Undersensing		
6799	2022	All other reportable incidents	47263	Cardiac resynchronization therapy implantable pacemaker	CT1684	Cardiac pacemakers and associated devices	E0109	Convulsion/Seizure	A070101	Failure to Capture	19/05/2022	19/05/2022
									A070909	Oversensing	19/05/2022	19/05/2022
									A0712	Pacing Problem	19/05/2022	19/05/2022
									A1107	Data Problem	19/05/2022	19/05/2022
							E0112	Dizziness	A070101	Failure to Capture	19/05/2022	19/05/2022
									A070909	Oversensing	19/05/2022	19/05/2022
									A0712	Pacing Problem	19/05/2022	19/05/2022
									A1107	Data Problem	19/05/2022	19/05/2022
							E060101	Asystole	A070101	Failure to Capture	19/05/2022	19/05/2022
									A070909	Oversensing	19/05/2022	19/05/2022
									A0712	Pacing Problem	19/05/2022	19/05/2022
									A1107	Data Problem	19/05/2022	19/05/2022
							E2403	No Clinical Signs, Symptoms or Conditions	A070101	Failure to Capture	19/05/2022	19/05/2022
									A070909	Oversensing	19/05/2022	19/05/2022
									A0712	Pacing Problem	19/05/2022	19/05/2022
									A1107	Data Problem	19/05/2022	19/05/2022
							F08	Hospitalization or Prolonged Hospitalization	A070101	Failure to Capture	19/05/2022	19/05/2022
									A070909	Oversensing	19/05/2022	19/05/2022
									A0712	Pacing Problem	19/05/2022	19/05/2022
									A1107	Data Problem	19/05/2022	19/05/2022
							F1905	Device Revision or Replacement	A070101	Failure to Capture	19/05/2022	19/05/2022
									A070909	Oversensing	19/05/2022	19/05/2022
									A0712	Pacing Problem	19/05/2022	19/05/2022
									A1107	Data Problem	19/05/2022	19/05/2022
					CT1685	Cardiac pacemakers	E0109	Convulsion/Seizure	A070101	Failure to Capture	19/05/2022	19/05/2022
									A070909	Oversensing	19/05/2022	19/05/2022
									A0712	Pacing Problem	19/05/2022	19/05/2022
									A1107	Data Problem	19/05/2022	19/05/2022
							E0112	Dizziness	A070101	Failure to Capture	19/05/2022	19/05/2022
									A070909	Oversensing	19/05/2022	19/05/2022
									A0712	Pacing Problem	19/05/2022	19/05/2022
									A1107	Data Problem	19/05/2022	19/05/2022
							E060101	Asystole	A070101	Failure to Capture	19/05/2022	19/05/2022
									A070909	Oversensing	19/05/2022	19/05/2022
									A0712	Pacing Problem	19/05/2022	19/05/2022
									A1107	Data Problem	19/05/2022	19/05/2022
							E2403	No Clinical Signs, Symptoms or Conditions	A070101	Failure to Capture	19/05/2022	19/05/2022
									A070909	Oversensing	19/05/2022	19/05/2022
									A0712	Pacing Problem	19/05/2022	19/05/2022
									A1107	Data Problem	19/05/2022	19/05/2022
							F08	Hospitalization or Prolonged Hospitalization	A070101	Failure to Capture	19/05/2022	19/05/2022
									A070909	Oversensing	19/05/2022	19/05/2022
									A0712	Pacing Problem	19/05/2022	19/05/2022
									A1107	Data Problem	19/05/2022	19/05/2022
							F1905	Device Revision or Replacement	A070101	Failure to Capture	19/05/2022	19/05/2022
									A070909	Oversensing	19/05/2022	19/05/2022
									A0712	Pacing Problem	19/05/2022	19/05/2022
									A1107	Data Problem	19/05/2022	19/05/2022
					CT257	Implantable cardiac pacemakers	E0109	Convulsion/Seizure	A070101	Failure to Capture	19/05/2022	19/05/2022
									A070909	Oversensing	19/05/2022	19/05/2022
									A0712	Pacing Problem	19/05/2022	19/05/2022
									A1107	Data Problem	19/05/2022	19/05/2022
							E0112	Dizziness	A070101	Failure to Capture	19/05/2022	19/05/2022
									A070909	Oversensing	19/05/2022	19/05/2022
									A0712	Pacing Problem	19/05/2022	19/05/2022
									A1107	Data Problem	19/05/2022	19/05/2022
							E060101	Asystole	A070101	Failure to Capture	19/05/2022	19/05/2022
									A070909	Oversensing	19/05/2022	19/05/2022
									A0712	Pacing Problem	19/05/2022	19/05/2022

								A1107	Data Problem	19/05/2022	19/05/2022
						E2403	No Clinical Signs, Symptoms or Conditions	A070101	Failure to Capture	19/05/2022	19/05/2022
								A070909	Oversensing	19/05/2022	19/05/2022
								A0712	Pacing Problem	19/05/2022	19/05/2022
								A1107	Data Problem	19/05/2022	19/05/2022
						F08	Hospitalization or Prolonged Hospitalization	A070101	Failure to Capture	19/05/2022	19/05/2022
								A070909	Oversensing	19/05/2022	19/05/2022
								A0712	Pacing Problem	19/05/2022	19/05/2022
								A1107	Data Problem	19/05/2022	19/05/2022
						F1905	Device Revision or Replacement	A070101	Failure to Capture	19/05/2022	19/05/2022
								A070909	Oversensing	19/05/2022	19/05/2022
								A0712	Pacing Problem	19/05/2022	19/05/2022
								A1107	Data Problem	19/05/2022	19/05/2022
					CT2866	Implantable cardiac electrical stimulators	E0109	Convulsion/Seizure	A070101	Failure to Capture	19/05/2022	19/05/2022
									A070909	Oversensing	19/05/2022	19/05/2022
									A0712	Pacing Problem	19/05/2022	19/05/2022
									A1107	Data Problem	19/05/2022	19/05/2022
							E0112	Dizziness	A070101	Failure to Capture	19/05/2022	19/05/2022
									A070909	Oversensing	19/05/2022	19/05/2022
									A0712	Pacing Problem	19/05/2022	19/05/2022
									A1107	Data Problem	19/05/2022	19/05/2022
							E060101	Asystole	A070101	Failure to Capture	19/05/2022	19/05/2022
									A070909	Oversensing	19/05/2022	19/05/2022
									A0712	Pacing Problem	19/05/2022	19/05/2022
									A1107	Data Problem	19/05/2022	19/05/2022
							E2403	No Clinical Signs, Symptoms or Conditions	A070101	Failure to Capture	19/05/2022	19/05/2022
									A070909	Oversensing	19/05/2022	19/05/2022
									A0712	Pacing Problem	19/05/2022	19/05/2022
									A1107	Data Problem	19/05/2022	19/05/2022
							F08	Hospitalization or Prolonged Hospitalization	A070101	Failure to Capture	19/05/2022	19/05/2022
									A070909	Oversensing	19/05/2022	19/05/2022
									A0712	Pacing Problem	19/05/2022	19/05/2022
									A1107	Data Problem	19/05/2022	19/05/2022
							F1905	Device Revision or Replacement	A070101	Failure to Capture	19/05/2022	19/05/2022
									A070909	Oversensing	19/05/2022	19/05/2022
									A0712	Pacing Problem	19/05/2022	19/05/2022
									A1107	Data Problem	19/05/2022	19/05/2022
6800	2022	All other reportable incidents	47263	Cardiac resynchronization therapy implantable pacemaker	CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A1107	Data Problem	20/05/2022	20/05/2022
							F08	Hospitalization or Prolonged Hospitalization	A1107	Data Problem	20/05/2022	20/05/2022
							F1905	Device Revision or Replacement	A1107	Data Problem	20/05/2022	20/05/2022
					CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A1107	Data Problem	20/05/2022	20/05/2022
							F08	Hospitalization or Prolonged Hospitalization	A1107	Data Problem	20/05/2022	20/05/2022
							F1905	Device Revision or Replacement	A1107	Data Problem	20/05/2022	20/05/2022
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A1107	Data Problem	20/05/2022	20/05/2022
							F08	Hospitalization or Prolonged Hospitalization	A1107	Data Problem	20/05/2022	20/05/2022
							F1905	Device Revision or Replacement	A1107	Data Problem	20/05/2022	20/05/2022
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A1107	Data Problem	20/05/2022	20/05/2022
							F08	Hospitalization or Prolonged Hospitalization	A1107	Data Problem	20/05/2022	20/05/2022
							F1905	Device Revision or Replacement	A1107	Data Problem	20/05/2022	20/05/2022
6801	2021	All other reportable incidents	35852	Single-chamber implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem	25/11/2021	25/11/2021
									A070504	Premature Discharge of Battery	25/11/2021	25/11/2021
							F08	Hospitalization or Prolonged Hospitalization	A0705	Battery Problem	25/11/2021	25/11/2021
									A070504	Premature Discharge of Battery	25/11/2021	25/11/2021
							F1905	Device Revision or Replacement	A0705	Battery Problem	25/11/2021	25/11/2021
									A070504	Premature Discharge of Battery	25/11/2021	25/11/2021
6802	2021	All other reportable incidents	47270	Cardiac resynchronization therapy implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E0601	Arrhythmia	A071202	Inaccurate Synchronization		
							E2403	No Clinical Signs, Symptoms or Conditions	A071202	Inaccurate Synchronization		
							F15	Recognized Device or Procedural Complication	A071202	Inaccurate Synchronization		
6803	2021	All other reportable incidents	47265	Dual-chamber implantable pacemaker, rate-responsive	CT141	Electrodes/Leads and associated devices	E011903	Syncope/Fainting	A0908	Incorrect, Inadequate or Imprecise Result or Readings	04/10/2021	04/10/2021
							E2403	No Clinical Signs, Symptoms or Conditions	A0908	Incorrect, Inadequate or Imprecise Result or Readings	04/10/2021	04/10/2021
							F08	Hospitalization or Prolonged Hospitalization	A0908	Incorrect, Inadequate or Imprecise Result or Readings	04/10/2021	04/10/2021
							F1905	Device Revision or Replacement	A0908	Incorrect, Inadequate or Imprecise Result or Readings	04/10/2021	04/10/2021
					CT145	Cardiovascular devices	E011903	Syncope/Fainting	A0908	Incorrect, Inadequate or Imprecise Result or Readings	04/10/2021	04/10/2021
							E2403	No Clinical Signs, Symptoms or Conditions	A0908	Incorrect, Inadequate or Imprecise Result or Readings	04/10/2021	04/10/2021
							F08	Hospitalization or Prolonged Hospitalization	A0908	Incorrect, Inadequate or Imprecise Result or Readings	04/10/2021	04/10/2021
							F1905	Device Revision or Replacement	A0908	Incorrect, Inadequate or Imprecise Result or Readings	04/10/2021	04/10/2021
					CT1684	Cardiac pacemakers and associated devices	E011903	Syncope/Fainting	A0908	Incorrect, Inadequate or Imprecise Result or Readings	04/10/2021	04/10/2021
							E2403	No Clinical Signs, Symptoms or Conditions	A0908	Incorrect, Inadequate or Imprecise Result or Readings	04/10/2021	04/10/2021
							F08	Hospitalization or Prolonged Hospitalization	A0908	Incorrect, Inadequate or Imprecise Result or Readings	04/10/2021	04/10/2021
							F1905	Device Revision or Replacement	A0908	Incorrect, Inadequate or Imprecise Result or Readings	04/10/2021	04/10/2021
					CT1685	Cardiac pacemakers	E011903	Syncope/Fainting	A0908	Incorrect, Inadequate or Imprecise Result or Readings	04/10/2021	04/10/2021
							E2403	No Clinical Signs, Symptoms or Conditions	A0908	Incorrect, Inadequate or Imprecise Result or Readings	04/10/2021	04/10/2021
							F08	Hospitalization or Prolonged Hospitalization	A0908	Incorrect, Inadequate or Imprecise Result or Readings	04/10/2021	04/10/2021
							F1905	Device Revision or Replacement	A0908	Incorrect, Inadequate or Imprecise Result or Readings	04/10/2021	04/10/2021
					CT233	Surgical	E011903	Syncope/Fainting	A0908	Incorrect, Inadequate or Imprecise Result or Readings	04/10/2021	04/10/2021

							E2403	No Clinical Signs, Symptoms or Conditions	A0908	Incorrect, Inadequate or Imprecise Result or Readings	04/10/2021	04/10/2021
							F08	Hospitalization or Prolonged Hospitalization	A0908	Incorrect, Inadequate or Imprecise Result or Readings	04/10/2021	04/10/2021
							F1905	Device Revision or Replacement	A0908	Incorrect, Inadequate or Imprecise Result or Readings	04/10/2021	04/10/2021
					CT2406	Implantable	E011903	Syncope/Fainting	A0908	Incorrect, Inadequate or Imprecise Result or Readings	04/10/2021	04/10/2021
							E2403	No Clinical Signs, Symptoms or Conditions	A0908	Incorrect, Inadequate or Imprecise Result or Readings	04/10/2021	04/10/2021
							F08	Hospitalization or Prolonged Hospitalization	A0908	Incorrect, Inadequate or Imprecise Result or Readings	04/10/2021	04/10/2021
							F1905	Device Revision or Replacement	A0908	Incorrect, Inadequate or Imprecise Result or Readings	04/10/2021	04/10/2021
					CT2407	Active implantable	E011903	Syncope/Fainting	A0908	Incorrect, Inadequate or Imprecise Result or Readings	04/10/2021	04/10/2021
							E2403	No Clinical Signs, Symptoms or Conditions	A0908	Incorrect, Inadequate or Imprecise Result or Readings	04/10/2021	04/10/2021
							F08	Hospitalization or Prolonged Hospitalization	A0908	Incorrect, Inadequate or Imprecise Result or Readings	04/10/2021	04/10/2021
							F1905	Device Revision or Replacement	A0908	Incorrect, Inadequate or Imprecise Result or Readings	04/10/2021	04/10/2021
					CT257	Implantable cardiac pacemakers	E011903	Syncope/Fainting	A0908	Incorrect, Inadequate or Imprecise Result or Readings	04/10/2021	04/10/2021
							E2403	No Clinical Signs, Symptoms or Conditions	A0908	Incorrect, Inadequate or Imprecise Result or Readings	04/10/2021	04/10/2021
							F08	Hospitalization or Prolonged Hospitalization	A0908	Incorrect, Inadequate or Imprecise Result or Readings	04/10/2021	04/10/2021
							F1905	Device Revision or Replacement	A0908	Incorrect, Inadequate or Imprecise Result or Readings	04/10/2021	04/10/2021
					CT2727	Active implant devices/ancillaries	E011903	Syncope/Fainting	A0908	Incorrect, Inadequate or Imprecise Result or Readings	04/10/2021	04/10/2021
							E2403	No Clinical Signs, Symptoms or Conditions	A0908	Incorrect, Inadequate or Imprecise Result or Readings	04/10/2021	04/10/2021
							F08	Hospitalization or Prolonged Hospitalization	A0908	Incorrect, Inadequate or Imprecise Result or Readings	04/10/2021	04/10/2021
							F1905	Device Revision or Replacement	A0908	Incorrect, Inadequate or Imprecise Result or Readings	04/10/2021	04/10/2021
					CT2750	Implantable cardiac electrical stimulators and associated devices	E011903	Syncope/Fainting	A0908	Incorrect, Inadequate or Imprecise Result or Readings	04/10/2021	04/10/2021
							E2403	No Clinical Signs, Symptoms or Conditions	A0908	Incorrect, Inadequate or Imprecise Result or Readings	04/10/2021	04/10/2021
							F08	Hospitalization or Prolonged Hospitalization	A0908	Incorrect, Inadequate or Imprecise Result or Readings	04/10/2021	04/10/2021
							F1905	Device Revision or Replacement	A0908	Incorrect, Inadequate or Imprecise Result or Readings	04/10/2021	04/10/2021
					CT2866	Implantable cardiac electrical stimulators	E011903	Syncope/Fainting	A0908	Incorrect, Inadequate or Imprecise Result or Readings	04/10/2021	04/10/2021
							E2403	No Clinical Signs, Symptoms or Conditions	A0908	Incorrect, Inadequate or Imprecise Result or Readings	04/10/2021	04/10/2021
							F08	Hospitalization or Prolonged Hospitalization	A0908	Incorrect, Inadequate or Imprecise Result or Readings	04/10/2021	04/10/2021
							F1905	Device Revision or Replacement	A0908	Incorrect, Inadequate or Imprecise Result or Readings	04/10/2021	04/10/2021
					CT2985	Active	E011903	Syncope/Fainting	A0908	Incorrect, Inadequate or Imprecise Result or Readings	04/10/2021	04/10/2021
							E2403	No Clinical Signs, Symptoms or Conditions	A0908	Incorrect, Inadequate or Imprecise Result or Readings	04/10/2021	04/10/2021
							F08	Hospitalization or Prolonged Hospitalization	A0908	Incorrect, Inadequate or Imprecise Result or Readings	04/10/2021	04/10/2021
							F1905	Device Revision or Replacement	A0908	Incorrect, Inadequate or Imprecise Result or Readings	04/10/2021	04/10/2021
					CT3001	Implantable metal	E011903	Syncope/Fainting	A0908	Incorrect, Inadequate or Imprecise Result or Readings	04/10/2021	04/10/2021
							E2403	No Clinical Signs, Symptoms or Conditions	A0908	Incorrect, Inadequate or Imprecise Result or Readings	04/10/2021	04/10/2021
							F08	Hospitalization or Prolonged Hospitalization	A0908	Incorrect, Inadequate or Imprecise Result or Readings	04/10/2021	04/10/2021
							F1905	Device Revision or Replacement	A0908	Incorrect, Inadequate or Imprecise Result or Readings	04/10/2021	04/10/2021
					CT3002	Implantable synthetic polymer	E011903	Syncope/Fainting	A0908	Incorrect, Inadequate or Imprecise Result or Readings	04/10/2021	04/10/2021
							E2403	No Clinical Signs, Symptoms or Conditions	A0908	Incorrect, Inadequate or Imprecise Result or Readings	04/10/2021	04/10/2021
							F08	Hospitalization or Prolonged Hospitalization	A0908	Incorrect, Inadequate or Imprecise Result or Readings	04/10/2021	04/10/2021
							F1905	Device Revision or Replacement	A0908	Incorrect, Inadequate or Imprecise Result or Readings	04/10/2021	04/10/2021
					CT3006	Implantable inorganic material	E011903	Syncope/Fainting	A0908	Incorrect, Inadequate or Imprecise Result or Readings	04/10/2021	04/10/2021
							E2403	No Clinical Signs, Symptoms or Conditions	A0908	Incorrect, Inadequate or Imprecise Result or Readings	04/10/2021	04/10/2021
							F08	Hospitalization or Prolonged Hospitalization	A0908	Incorrect, Inadequate or Imprecise Result or Readings	04/10/2021	04/10/2021
							F1905	Device Revision or Replacement	A0908	Incorrect, Inadequate or Imprecise Result or Readings	04/10/2021	04/10/2021
					CT314	Electrically powered	E011903	Syncope/Fainting	A0908	Incorrect, Inadequate or Imprecise Result or Readings	04/10/2021	04/10/2021
							E2403	No Clinical Signs, Symptoms or Conditions	A0908	Incorrect, Inadequate or Imprecise Result or Readings	04/10/2021	04/10/2021
							F08	Hospitalization or Prolonged Hospitalization	A0908	Incorrect, Inadequate or Imprecise Result or Readings	04/10/2021	04/10/2021
							F1905	Device Revision or Replacement	A0908	Incorrect, Inadequate or Imprecise Result or Readings	04/10/2021	04/10/2021
					CT321	Long-term surgical invasive	E011903	Syncope/Fainting	A0908	Incorrect, Inadequate or Imprecise Result or Readings	04/10/2021	04/10/2021
							E2403	No Clinical Signs, Symptoms or Conditions	A0908	Incorrect, Inadequate or Imprecise Result or Readings	04/10/2021	04/10/2021
							F08	Hospitalization or Prolonged Hospitalization	A0908	Incorrect, Inadequate or Imprecise Result or Readings	04/10/2021	04/10/2021

							F1905	Device Revision or Replacement	A0908	Incorrect, Inadequate or Imprecise Result or Readings	04/10/2021	04/10/2021
					CT449	Generators and associated devices	E011903	Syncope/Fainting	A0908	Incorrect, Inadequate or Imprecise Result or Readings	04/10/2021	04/10/2021
							E2403	No Clinical Signs, Symptoms or Conditions	A0908	Incorrect, Inadequate or Imprecise Result or Readings	04/10/2021	04/10/2021
							F08	Hospitalization or Prolonged Hospitalization	A0908	Incorrect, Inadequate or Imprecise Result or Readings	04/10/2021	04/10/2021
							F1905	Device Revision or Replacement	A0908	Incorrect, Inadequate or Imprecise Result or Readings	04/10/2021	04/10/2021
					CT981	Single use	E011903	Syncope/Fainting	A0908	Incorrect, Inadequate or Imprecise Result or Readings	04/10/2021	04/10/2021
							E2403	No Clinical Signs, Symptoms or Conditions	A0908	Incorrect, Inadequate or Imprecise Result or Readings	04/10/2021	04/10/2021
							F08	Hospitalization or Prolonged Hospitalization	A0908	Incorrect, Inadequate or Imprecise Result or Readings	04/10/2021	04/10/2021
							F1905	Device Revision or Replacement	A0908	Incorrect, Inadequate or Imprecise Result or Readings	04/10/2021	04/10/2021
					CT983	Surgical invasive	E011903	Syncope/Fainting	A0908	Incorrect, Inadequate or Imprecise Result or Readings	04/10/2021	04/10/2021
							E2403	No Clinical Signs, Symptoms or Conditions	A0908	Incorrect, Inadequate or Imprecise Result or Readings	04/10/2021	04/10/2021
							F08	Hospitalization or Prolonged Hospitalization	A0908	Incorrect, Inadequate or Imprecise Result or Readings	04/10/2021	04/10/2021
							F1905	Device Revision or Replacement	A0908	Incorrect, Inadequate or Imprecise Result or Readings	04/10/2021	04/10/2021
6804	2021	All other reportable incidents	35223	Endocardial/interventricular septal pacing lead	CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A0401	Break		
									A07	Electrical /Electronic Property Problem		
									A070901	Decreased Sensitivity		
									A070909	Oversensing		
							F26	No Health Consequences or Impact	A0401	Break		
									A07	Electrical /Electronic Property Problem		
									A070901	Decreased Sensitivity		
									A070909	Oversensing		
6805	2021	All other reportable incidents	37265	Dual-chamber implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A0712	Pacing Problem		
									A071302	Inappropriate/Inadequate Shock/Stimulation		
							F26	No Health Consequences or Impact	A0712	Pacing Problem		
									A071302	Inappropriate/Inadequate Shock/Stimulation		
6806	2021	All other reportable incidents	36241	Pacemaker/defibrillator lead	CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A072201	High Impedance		
							F26	No Health Consequences or Impact	A072201	High Impedance		
6807	2021	All other reportable incidents	35852	Single-chamber implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem	09/11/2021	09/11/2021
									A070504	Premature Discharge of Battery	09/11/2021	09/11/2021
							F08	Hospitalization or Prolonged Hospitalization	A0705	Battery Problem	09/11/2021	09/11/2021
									A070504	Premature Discharge of Battery	09/11/2021	09/11/2021
							F1905	Device Revision or Replacement	A0705	Battery Problem	09/11/2021	09/11/2021
									A070504	Premature Discharge of Battery	09/11/2021	09/11/2021
6808	2021	All other reportable incidents	47263	Cardiac resynchronization therapy implantable pacemaker	CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem		
							F26	No Health Consequences or Impact	A0705	Battery Problem		
					CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem		
							F26	No Health Consequences or Impact	A0705	Battery Problem		
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem		
							F26	No Health Consequences or Impact	A0705	Battery Problem		
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem		
							F26	No Health Consequences or Impact	A0705	Battery Problem		
6809	2021	All other reportable incidents	36241	Pacemaker/defibrillator lead	CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070909	Oversensing		
							F23	Unexpected Medical Intervention	A070909	Oversensing		
6810	2021	All other reportable incidents	37265	Dual-chamber implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E0601	Arrhythmia	A0712	Pacing Problem		
									A071202	Inaccurate Synchronization		
							E2403	No Clinical Signs, Symptoms or Conditions	A0712	Pacing Problem		
									A071202	Inaccurate Synchronization		
							F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A0712	Pacing Problem		
									A071202	Inaccurate Synchronization		
6811	2022	All other reportable incidents	60789	Intracardiac pacemaker	CT130	Catheters and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070102	High Capture Threshold		
									A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070102	High Capture Threshold		
									A070504	Premature Discharge of Battery		
					CT145	Cardiovascular devices	E2403	No Clinical Signs, Symptoms or Conditions	A070102	High Capture Threshold		
									A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070102	High Capture Threshold		
									A070504	Premature Discharge of Battery		
					CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070102	High Capture Threshold		
									A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070102	High Capture Threshold		
									A070504	Premature Discharge of Battery		
					CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A070102	High Capture Threshold		
									A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070102	High Capture Threshold		
									A070504	Premature Discharge of Battery		
					CT233	Surgical	E2403	No Clinical Signs, Symptoms or Conditions	A070102	High Capture Threshold		
									A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070102	High Capture Threshold		
									A070504	Premature Discharge of Battery		
					CT2406	Implantable	E2403	No Clinical Signs, Symptoms or Conditions	A070102	High Capture Threshold		
									A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070102	High Capture Threshold		
									A070504	Premature Discharge of Battery		
					CT2407	Active implantable	E2403	No Clinical Signs, Symptoms or Conditions	A070102	High Capture Threshold		
									A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070102	High Capture Threshold		
									A070504	Premature Discharge of Battery		
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A070102	High Capture Threshold		

								A070504	Premature Discharge of Battery		
						F26	No Health Consequences or Impact	A070102	High Capture Threshold		
								A070504	Premature Discharge of Battery		
					CT2727	Active implant devices/ancillaries	E2403	No Clinical Signs, Symptoms or Conditions	A070102	High Capture Threshold		
									A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070102	High Capture Threshold		
									A070504	Premature Discharge of Battery		
					CT2750	Implantable cardiac electrical stimulators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070102	High Capture Threshold		
									A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070102	High Capture Threshold		
									A070504	Premature Discharge of Battery		
					CT2818	Catheterization/Cannulation/Fluid path devices	E2403	No Clinical Signs, Symptoms or Conditions	A070102	High Capture Threshold		
									A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070102	High Capture Threshold		
									A070504	Premature Discharge of Battery		
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A070102	High Capture Threshold		
									A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070102	High Capture Threshold		
									A070504	Premature Discharge of Battery		
					CT2985	Active	E2403	No Clinical Signs, Symptoms or Conditions	A070102	High Capture Threshold		
									A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070102	High Capture Threshold		
									A070504	Premature Discharge of Battery		
					CT3001	Implantable metal	E2403	No Clinical Signs, Symptoms or Conditions	A070102	High Capture Threshold		
									A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070102	High Capture Threshold		
									A070504	Premature Discharge of Battery		
					CT3002	Implantable synthetic polymer	E2403	No Clinical Signs, Symptoms or Conditions	A070102	High Capture Threshold		
									A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070102	High Capture Threshold		
									A070504	Premature Discharge of Battery		
					CT3006	Implantable inorganic material	E2403	No Clinical Signs, Symptoms or Conditions	A070102	High Capture Threshold		
									A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070102	High Capture Threshold		
									A070504	Premature Discharge of Battery		
					CT314	Electrically powered	E2403	No Clinical Signs, Symptoms or Conditions	A070102	High Capture Threshold		
									A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070102	High Capture Threshold		
									A070504	Premature Discharge of Battery		
					CT321	Long-term surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A070102	High Capture Threshold		
									A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070102	High Capture Threshold		
									A070504	Premature Discharge of Battery		
					CT981	Single use	E2403	No Clinical Signs, Symptoms or Conditions	A070102	High Capture Threshold		
									A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070102	High Capture Threshold		
									A070504	Premature Discharge of Battery		
					CT983	Surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A070102	High Capture Threshold		
									A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070102	High Capture Threshold		
									A070504	Premature Discharge of Battery		
6812	2022	All other reportable incidents	35852	Single-chamber implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem	28/06/2022	29/06/2022
									A070504	Premature Discharge of Battery	28/06/2022	29/06/2022
							F08	Hospitalization or Prolonged Hospitalization	A0705	Battery Problem	28/06/2022	29/06/2022
									A070504	Premature Discharge of Battery	28/06/2022	29/06/2022
							F1905	Device Revision or Replacement	A0705	Battery Problem	28/06/2022	29/06/2022
									A070504	Premature Discharge of Battery	28/06/2022	29/06/2022
6813	2022	All other reportable incidents	47265	Dual-chamber implantable pacemaker, rate-responsive	CT141	Electrodes/Leads and associated devices	E060101	Asystole	A070909	Oversensing		
									A0712	Pacing Problem		
							E2403	No Clinical Signs, Symptoms or Conditions	A070909	Oversensing		
									A0712	Pacing Problem		
							F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A070909	Oversensing		
									A0712	Pacing Problem		
							F1001	Absence of Treatment	A070909	Oversensing		
									A0712	Pacing Problem		
					CT145	Cardiovascular devices	E060101	Asystole	A070909	Oversensing		
									A0712	Pacing Problem		
							E2403	No Clinical Signs, Symptoms or Conditions	A070909	Oversensing		
									A0712	Pacing Problem		
							F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A070909	Oversensing		
									A0712	Pacing Problem		
							F1001	Absence of Treatment	A070909	Oversensing		
									A0712	Pacing Problem		
					CT1684	Cardiac pacemakers and associated devices	E060101	Asystole	A070909	Oversensing		
									A0712	Pacing Problem		
							E2403	No Clinical Signs, Symptoms or Conditions	A070909	Oversensing		
									A0712	Pacing Problem		
							F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A070909	Oversensing		
									A0712	Pacing Problem		
							F1001	Absence of Treatment	A070909	Oversensing		
									A0712	Pacing Problem		
					CT1685	Cardiac pacemakers	E060101	Asystole	A070909	Oversensing		
									A0712	Pacing Problem		
							E2403	No Clinical Signs, Symptoms or Conditions	A070909	Oversensing		
									A0712	Pacing Problem		
							F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A070909	Oversensing		
									A0712	Pacing Problem		
							F1001	Absence of Treatment	A070909	Oversensing		

									A0712	Pacing Problem		
					CT233	Surgical	E060101	Asystole	A070909	Oversensing		
									A0712	Pacing Problem		
							E2403	No Clinical Signs, Symptoms or Conditions	A070909	Oversensing		
									A0712	Pacing Problem		
							F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A070909	Oversensing		
									A0712	Pacing Problem		
							F1001	Absence of Treatment	A070909	Oversensing		
									A0712	Pacing Problem		
					CT2406	Implantable	E060101	Asystole	A070909	Oversensing		
									A0712	Pacing Problem		
							E2403	No Clinical Signs, Symptoms or Conditions	A070909	Oversensing		
									A0712	Pacing Problem		
							F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A070909	Oversensing		
									A0712	Pacing Problem		
							F1001	Absence of Treatment	A070909	Oversensing		
									A0712	Pacing Problem		
					CT2407	Active implantable	E060101	Asystole	A070909	Oversensing		
									A0712	Pacing Problem		
							E2403	No Clinical Signs, Symptoms or Conditions	A070909	Oversensing		
									A0712	Pacing Problem		
							F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A070909	Oversensing		
									A0712	Pacing Problem		
							F1001	Absence of Treatment	A070909	Oversensing		
									A0712	Pacing Problem		
					CT257	Implantable cardiac pacemakers	E060101	Asystole	A070909	Oversensing		
									A0712	Pacing Problem		
							E2403	No Clinical Signs, Symptoms or Conditions	A070909	Oversensing		
									A0712	Pacing Problem		
							F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A070909	Oversensing		
									A0712	Pacing Problem		
							F1001	Absence of Treatment	A070909	Oversensing		
									A0712	Pacing Problem		
					CT2727	Active implant devices/ancillaries	E060101	Asystole	A070909	Oversensing		
									A0712	Pacing Problem		
							E2403	No Clinical Signs, Symptoms or Conditions	A070909	Oversensing		
									A0712	Pacing Problem		
							F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A070909	Oversensing		
									A0712	Pacing Problem		
							F1001	Absence of Treatment	A070909	Oversensing		
									A0712	Pacing Problem		
					CT2750	Implantable cardiac electrical stimulators and associated devices	E060101	Asystole	A070909	Oversensing		
									A0712	Pacing Problem		
							E2403	No Clinical Signs, Symptoms or Conditions	A070909	Oversensing		
									A0712	Pacing Problem		
							F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A070909	Oversensing		
									A0712	Pacing Problem		
							F1001	Absence of Treatment	A070909	Oversensing		
									A0712	Pacing Problem		
					CT2866	Implantable cardiac electrical stimulators	E060101	Asystole	A070909	Oversensing		
									A0712	Pacing Problem		
							E2403	No Clinical Signs, Symptoms or Conditions	A070909	Oversensing		
									A0712	Pacing Problem		
							F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A070909	Oversensing		
									A0712	Pacing Problem		
							F1001	Absence of Treatment	A070909	Oversensing		
									A0712	Pacing Problem		
					CT2985	Active	E060101	Asystole	A070909	Oversensing		
									A0712	Pacing Problem		
							E2403	No Clinical Signs, Symptoms or Conditions	A070909	Oversensing		
									A0712	Pacing Problem		
							F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A070909	Oversensing		
									A0712	Pacing Problem		
							F1001	Absence of Treatment	A070909	Oversensing		
									A0712	Pacing Problem		
					CT3001	Implantable metal	E060101	Asystole	A070909	Oversensing		
									A0712	Pacing Problem		
							E2403	No Clinical Signs, Symptoms or Conditions	A070909	Oversensing		
									A0712	Pacing Problem		
							F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A070909	Oversensing		
									A0712	Pacing Problem		
							F1001	Absence of Treatment	A070909	Oversensing		
									A0712	Pacing Problem		
					CT3002	Implantable synthetic polymer	E060101	Asystole	A070909	Oversensing		
									A0712	Pacing Problem		
							E2403	No Clinical Signs, Symptoms or Conditions	A070909	Oversensing		
									A0712	Pacing Problem		
							F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A070909	Oversensing		
									A0712	Pacing Problem		
							F1001	Absence of Treatment	A070909	Oversensing		
									A0712	Pacing Problem		
					CT3006	Implantable inorganic material	E060101	Asystole	A070909	Oversensing		
									A0712	Pacing Problem		
							E2403	No Clinical Signs, Symptoms or Conditions	A070909	Oversensing		
									A0712	Pacing Problem		
							F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A070909	Oversensing		
									A0712	Pacing Problem		
							F1001	Absence of Treatment	A070909	Oversensing		

								A0712	Pacing Problem		
				CT314	Electrically powered	E060101	Asystole	A070909	Oversensing		
								A0712	Pacing Problem		
						E2403	No Clinical Signs, Symptoms or Conditions	A070909	Oversensing		
								A0712	Pacing Problem		
						F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A070909	Oversensing		
								A0712	Pacing Problem		
						F1001	Absence of Treatment	A070909	Oversensing		
								A0712	Pacing Problem		
				CT321	Long-term surgical invasive	E060101	Asystole	A070909	Oversensing		
								A0712	Pacing Problem		
						E2403	No Clinical Signs, Symptoms or Conditions	A070909	Oversensing		
								A0712	Pacing Problem		
						F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A070909	Oversensing		
								A0712	Pacing Problem		
						F1001	Absence of Treatment	A070909	Oversensing		
								A0712	Pacing Problem		
				CT449	Generators and associated devices	E060101	Asystole	A070909	Oversensing		
								A0712	Pacing Problem		
						E2403	No Clinical Signs, Symptoms or Conditions	A070909	Oversensing		
								A0712	Pacing Problem		
						F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A070909	Oversensing		
								A0712	Pacing Problem		
						F1001	Absence of Treatment	A070909	Oversensing		
								A0712	Pacing Problem		
				CT981	Single use	E060101	Asystole	A070909	Oversensing		
								A0712	Pacing Problem		
						E2403	No Clinical Signs, Symptoms or Conditions	A070909	Oversensing		
								A0712	Pacing Problem		
						F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A070909	Oversensing		
								A0712	Pacing Problem		
						F1001	Absence of Treatment	A070909	Oversensing		
								A0712	Pacing Problem		
				CT983	Surgical invasive	E060101	Asystole	A070909	Oversensing		
								A0712	Pacing Problem		
						E2403	No Clinical Signs, Symptoms or Conditions	A070909	Oversensing		
								A0712	Pacing Problem		
						F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A070909	Oversensing		
								A0712	Pacing Problem		
						F1001	Absence of Treatment	A070909	Oversensing		
								A0712	Pacing Problem		
6814	2022	All other reportable incidents	36241	Pacemaker/defibrillator lead	CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A040101	Fracture	06/12/2016	06/12/2016
									A070102	High Capture Threshold	06/12/2016	06/12/2016
							F08	Hospitalization or Prolonged Hospitalization	A040101	Fracture	06/12/2016	06/12/2016
									A070102	High Capture Threshold	06/12/2016	06/12/2016
							F1905	Device Revision or Replacement	A040101	Fracture	06/12/2016	06/12/2016
									A070102	High Capture Threshold	06/12/2016	06/12/2016
6815	2022	All other reportable incidents	47263	Cardiac resynchronization therapy implantable pacemaker	CT1684	Cardiac pacemakers and associated devices	E060101	Asystole	A070909	Oversensing		
									A0712	Pacing Problem		
									A0908	Incorrect, Inadequate or Imprecise Result or Readings		
							E2403	No Clinical Signs, Symptoms or Conditions	A070909	Oversensing		
									A0712	Pacing Problem		
									A0908	Incorrect, Inadequate or Imprecise Result or Readings		
							F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A070909	Oversensing		
									A0712	Pacing Problem		
									A0908	Incorrect, Inadequate or Imprecise Result or Readings		
							F1001	Absence of Treatment	A070909	Oversensing		
									A0712	Pacing Problem		
									A0908	Incorrect, Inadequate or Imprecise Result or Readings		
					CT1685	Cardiac pacemakers	E060101	Asystole	A070909	Oversensing		
									A0712	Pacing Problem		
									A0908	Incorrect, Inadequate or Imprecise Result or Readings		
							E2403	No Clinical Signs, Symptoms or Conditions	A070909	Oversensing		
									A0712	Pacing Problem		
									A0908	Incorrect, Inadequate or Imprecise Result or Readings		
							F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A070909	Oversensing		
									A0712	Pacing Problem		
									A0908	Incorrect, Inadequate or Imprecise Result or Readings		
							F1001	Absence of Treatment	A070909	Oversensing		
									A0712	Pacing Problem		
									A0908	Incorrect, Inadequate or Imprecise Result or Readings		
					CT257	Implantable cardiac pacemakers	E060101	Asystole	A070909	Oversensing		
									A0712	Pacing Problem		
									A0908	Incorrect, Inadequate or Imprecise Result or Readings		
							E2403	No Clinical Signs, Symptoms or Conditions	A070909	Oversensing		
									A0712	Pacing Problem		
									A0908	Incorrect, Inadequate or Imprecise Result or Readings		
							F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A070909	Oversensing		
									A0712	Pacing Problem		

								A0908	Incorrect, Inadequate or Imprecise Result or Readings		
						F1001	Absence of Treatment	A070909	Oversensing		
								A0712	Pacing Problem		
								A0908	Incorrect, Inadequate or Imprecise Result or Readings		
					CT2866	Implantable cardiac electrical stimulators	E060101	Asystole	A070909	Oversensing	
									A0712	Pacing Problem	
									A0908	Incorrect, Inadequate or Imprecise Result or Readings	
							E2403	No Clinical Signs, Symptoms or Conditions	A070909	Oversensing	
									A0712	Pacing Problem	
									A0908	Incorrect, Inadequate or Imprecise Result or Readings	
							F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A070909	Oversensing	
									A0712	Pacing Problem	
									A0908	Incorrect, Inadequate or Imprecise Result or Readings	
							F1001	Absence of Treatment	A070909	Oversensing	
									A0712	Pacing Problem	
									A0908	Incorrect, Inadequate or Imprecise Result or Readings	
6816	2022	All other reportable incidents	47265	Dual-chamber implantable pacemaker, rate-responsive	CT141	Electrodes/Leads and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem	
									A070504	Premature Discharge of Battery	
							F24	Insufficient Health Impact Information	A0705	Battery Problem	
									A070504	Premature Discharge of Battery	
					CT145	Cardiovascular devices	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem	
									A070504	Premature Discharge of Battery	
							F24	Insufficient Health Impact Information	A0705	Battery Problem	
									A070504	Premature Discharge of Battery	
					CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem	
									A070504	Premature Discharge of Battery	
							F24	Insufficient Health Impact Information	A0705	Battery Problem	
									A070504	Premature Discharge of Battery	
					CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem	
									A070504	Premature Discharge of Battery	
							F24	Insufficient Health Impact Information	A0705	Battery Problem	
									A070504	Premature Discharge of Battery	
					CT233	Surgical	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem	
									A070504	Premature Discharge of Battery	
							F24	Insufficient Health Impact Information	A0705	Battery Problem	
									A070504	Premature Discharge of Battery	
					CT2406	Implantable	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem	
									A070504	Premature Discharge of Battery	
							F24	Insufficient Health Impact Information	A0705	Battery Problem	
									A070504	Premature Discharge of Battery	
					CT2407	Active implantable	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem	
									A070504	Premature Discharge of Battery	
							F24	Insufficient Health Impact Information	A0705	Battery Problem	
									A070504	Premature Discharge of Battery	
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem	
									A070504	Premature Discharge of Battery	
							F24	Insufficient Health Impact Information	A0705	Battery Problem	
									A070504	Premature Discharge of Battery	
					CT2727	Active implant devices/ancillaries	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem	
									A070504	Premature Discharge of Battery	
							F24	Insufficient Health Impact Information	A0705	Battery Problem	
									A070504	Premature Discharge of Battery	
					CT2750	Implantable cardiac electrical stimulators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem	
									A070504	Premature Discharge of Battery	
							F24	Insufficient Health Impact Information	A0705	Battery Problem	
									A070504	Premature Discharge of Battery	
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem	
									A070504	Premature Discharge of Battery	
							F24	Insufficient Health Impact Information	A0705	Battery Problem	
									A070504	Premature Discharge of Battery	
					CT2985	Active	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem	
									A070504	Premature Discharge of Battery	
							F24	Insufficient Health Impact Information	A0705	Battery Problem	
									A070504	Premature Discharge of Battery	
					CT3001	Implantable metal	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem	
									A070504	Premature Discharge of Battery	
							F24	Insufficient Health Impact Information	A0705	Battery Problem	
									A070504	Premature Discharge of Battery	
					CT3002	Implantable synthetic polymer	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem	
									A070504	Premature Discharge of Battery	
							F24	Insufficient Health Impact Information	A0705	Battery Problem	
									A070504	Premature Discharge of Battery	
					CT3006	Implantable inorganic material	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem	
									A070504	Premature Discharge of Battery	
							F24	Insufficient Health Impact Information	A0705	Battery Problem	
									A070504	Premature Discharge of Battery	
					CT314	Electrically powered	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem	
									A070504	Premature Discharge of Battery	
							F24	Insufficient Health Impact Information	A0705	Battery Problem	
									A070504	Premature Discharge of Battery	
					CT321	Long-term surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem	
									A070504	Premature Discharge of Battery	
							F24	Insufficient Health Impact Information	A0705	Battery Problem	

									A070504	Premature Discharge of Battery		
					CT449	Generators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
							F24	Insufficient Health Impact Information	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
					CT981	Single use	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
							F24	Insufficient Health Impact Information	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
					CT983	Surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
							F24	Insufficient Health Impact Information	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
6817	2022	Unanticipated serious deterioration in state of health	17882	Physiologic-monitoring defibrillation system	CT1684	Cardiac pacemakers and associated devices	E2401	Insufficient Information	A10	Temperature Problem		
							F12	Serious Injury/ Illness/ Impairment	A10	Temperature Problem		
6818	2022	All other reportable incidents	47265	Dual-chamber implantable pacemaker, rate-responsive	CT141	Electrodes/Leads and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT145	Cardiovascular devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT233	Surgical	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT2406	Implantable	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT2407	Active implantable	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT2727	Active implant devices/ancillaries	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT2750	Implantable cardiac electrical stimulators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT2985	Active	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT3001	Implantable metal	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT3002	Implantable synthetic polymer	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT3006	Implantable inorganic material	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT314	Electrically powered	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT321	Long-term surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT449	Generators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT981	Single use	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT983	Surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
6819	2022	All other reportable incidents	47265	Dual-chamber implantable pacemaker, rate-responsive	CT141	Electrodes/Leads and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070909	Oversensing		
									A072201	High Impedance		
							F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A070909	Oversensing		
									A072201	High Impedance		
							F2203	Imaging Required	A070909	Oversensing		
									A072201	High Impedance		
					CT145	Cardiovascular devices	E2403	No Clinical Signs, Symptoms or Conditions	A070909	Oversensing		
									A072201	High Impedance		
							F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A070909	Oversensing		
									A072201	High Impedance		
							F2203	Imaging Required	A070909	Oversensing		
									A072201	High Impedance		
					CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070909	Oversensing		
									A072201	High Impedance		
							F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A070909	Oversensing		
									A072201	High Impedance		
							F2203	Imaging Required	A070909	Oversensing		
									A072201	High Impedance		
					CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A070909	Oversensing		
									A072201	High Impedance		
							F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A070909	Oversensing		
									A072201	High Impedance		
							F2203	Imaging Required	A070909	Oversensing		
									A072201	High Impedance		
					CT233	Surgical	E2403	No Clinical Signs, Symptoms or Conditions	A070909	Oversensing		
									A072201	High Impedance		
							F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A070909	Oversensing		
									A072201	High Impedance		
							F2203	Imaging Required	A070909	Oversensing		
									A072201	High Impedance		
					CT2406	Implantable	E2403	No Clinical Signs, Symptoms or Conditions	A070909	Oversensing		
									A072201	High Impedance		
							F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A070909	Oversensing		

									A072201	High Impedance		
							F2203	Imaging Required	A070909	Oversensing		
									A072201	High Impedance		
					CT2407	Active implantable	E2403	No Clinical Signs, Symptoms or Conditions	A070909	Oversensing		
									A072201	High Impedance		
							F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A070909	Oversensing		
									A072201	High Impedance		
							F2203	Imaging Required	A070909	Oversensing		
									A072201	High Impedance		
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A070909	Oversensing		
									A072201	High Impedance		
							F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A070909	Oversensing		
									A072201	High Impedance		
							F2203	Imaging Required	A070909	Oversensing		
									A072201	High Impedance		
					CT2727	Active implant devices/ancillaries	E2403	No Clinical Signs, Symptoms or Conditions	A070909	Oversensing		
									A072201	High Impedance		
							F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A070909	Oversensing		
									A072201	High Impedance		
							F2203	Imaging Required	A070909	Oversensing		
									A072201	High Impedance		
					CT2750	Implantable cardiac electrical stimulators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070909	Oversensing		
									A072201	High Impedance		
							F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A070909	Oversensing		
									A072201	High Impedance		
							F2203	Imaging Required	A070909	Oversensing		
									A072201	High Impedance		
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A070909	Oversensing		
									A072201	High Impedance		
							F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A070909	Oversensing		
									A072201	High Impedance		
							F2203	Imaging Required	A070909	Oversensing		
									A072201	High Impedance		
					CT2985	Active	E2403	No Clinical Signs, Symptoms or Conditions	A070909	Oversensing		
									A072201	High Impedance		
							F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A070909	Oversensing		
									A072201	High Impedance		
							F2203	Imaging Required	A070909	Oversensing		
									A072201	High Impedance		
					CT3001	Implantable metal	E2403	No Clinical Signs, Symptoms or Conditions	A070909	Oversensing		
									A072201	High Impedance		
							F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A070909	Oversensing		
									A072201	High Impedance		
							F2203	Imaging Required	A070909	Oversensing		
									A072201	High Impedance		
					CT3002	Implantable synthetic polymer	E2403	No Clinical Signs, Symptoms or Conditions	A070909	Oversensing		
									A072201	High Impedance		
							F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A070909	Oversensing		
									A072201	High Impedance		
							F2203	Imaging Required	A070909	Oversensing		
									A072201	High Impedance		
					CT3006	Implantable inorganic material	E2403	No Clinical Signs, Symptoms or Conditions	A070909	Oversensing		
									A072201	High Impedance		
							F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A070909	Oversensing		
									A072201	High Impedance		
							F2203	Imaging Required	A070909	Oversensing		
									A072201	High Impedance		
					CT314	Electrically powered	E2403	No Clinical Signs, Symptoms or Conditions	A070909	Oversensing		
									A072201	High Impedance		
							F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A070909	Oversensing		
									A072201	High Impedance		
							F2203	Imaging Required	A070909	Oversensing		
									A072201	High Impedance		
					CT321	Long-term surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A070909	Oversensing		
									A072201	High Impedance		
							F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A070909	Oversensing		
									A072201	High Impedance		
							F2203	Imaging Required	A070909	Oversensing		
									A072201	High Impedance		
					CT449	Generators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070909	Oversensing		
									A072201	High Impedance		
							F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A070909	Oversensing		
									A072201	High Impedance		
							F2203	Imaging Required	A070909	Oversensing		
									A072201	High Impedance		
					CT981	Single use	E2403	No Clinical Signs, Symptoms or Conditions	A070909	Oversensing		
									A072201	High Impedance		
							F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A070909	Oversensing		
									A072201	High Impedance		
							F2203	Imaging Required	A070909	Oversensing		
									A072201	High Impedance		
					CT983	Surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A070909	Oversensing		
									A072201	High Impedance		
							F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A070909	Oversensing		
									A072201	High Impedance		
							F2203	Imaging Required	A070909	Oversensing		
									A072201	High Impedance		
6820	2022	All other reportable incidents	36241	Pacemaker/defibrillator lead	CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070909	Oversensing		

									A090801	Signal Artifact/Noise		
							F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A070909	Oversensing		
									A090801	Signal Artifact/Noise		
6821	2021	All other reportable incidents	47265	Dual-chamber implantable pacemaker, rate-responsive	CT141	Electrodes/Leads and associated devices	E0717	Dyspnea	A0701	Capturing Problem	26/07/2021	26/07/2021
									A072201	High Impedance	26/07/2021	26/07/2021
									A090805	Incorrect Measurement	26/07/2021	26/07/2021
							E2403	No Clinical Signs, Symptoms or Conditions	A0701	Capturing Problem	26/07/2021	26/07/2021
									A072201	High Impedance	26/07/2021	26/07/2021
									A090805	Incorrect Measurement	26/07/2021	26/07/2021
							F1901	Additional Surgery	A0701	Capturing Problem	26/07/2021	26/07/2021
									A072201	High Impedance	26/07/2021	26/07/2021
									A090805	Incorrect Measurement	26/07/2021	26/07/2021
					CT145	Cardiovascular devices	E0717	Dyspnea	A0701	Capturing Problem	26/07/2021	26/07/2021
									A072201	High Impedance	26/07/2021	26/07/2021
									A090805	Incorrect Measurement	26/07/2021	26/07/2021
							E2403	No Clinical Signs, Symptoms or Conditions	A0701	Capturing Problem	26/07/2021	26/07/2021
									A072201	High Impedance	26/07/2021	26/07/2021
									A090805	Incorrect Measurement	26/07/2021	26/07/2021
							F1901	Additional Surgery	A0701	Capturing Problem	26/07/2021	26/07/2021
									A072201	High Impedance	26/07/2021	26/07/2021
									A090805	Incorrect Measurement	26/07/2021	26/07/2021
					CT1684	Cardiac pacemakers and associated devices	E0717	Dyspnea	A0701	Capturing Problem	26/07/2021	26/07/2021
									A072201	High Impedance	26/07/2021	26/07/2021
									A090805	Incorrect Measurement	26/07/2021	26/07/2021
							E2403	No Clinical Signs, Symptoms or Conditions	A0701	Capturing Problem	26/07/2021	26/07/2021
									A072201	High Impedance	26/07/2021	26/07/2021
									A090805	Incorrect Measurement	26/07/2021	26/07/2021
							F1901	Additional Surgery	A0701	Capturing Problem	26/07/2021	26/07/2021
									A072201	High Impedance	26/07/2021	26/07/2021
									A090805	Incorrect Measurement	26/07/2021	26/07/2021
					CT1685	Cardiac pacemakers	E0717	Dyspnea	A0701	Capturing Problem	26/07/2021	26/07/2021
									A072201	High Impedance	26/07/2021	26/07/2021
									A090805	Incorrect Measurement	26/07/2021	26/07/2021
							E2403	No Clinical Signs, Symptoms or Conditions	A0701	Capturing Problem	26/07/2021	26/07/2021
									A072201	High Impedance	26/07/2021	26/07/2021
									A090805	Incorrect Measurement	26/07/2021	26/07/2021
							F1901	Additional Surgery	A0701	Capturing Problem	26/07/2021	26/07/2021
									A072201	High Impedance	26/07/2021	26/07/2021
									A090805	Incorrect Measurement	26/07/2021	26/07/2021
					CT233	Surgical	E0717	Dyspnea	A0701	Capturing Problem	26/07/2021	26/07/2021
									A072201	High Impedance	26/07/2021	26/07/2021
									A090805	Incorrect Measurement	26/07/2021	26/07/2021
							E2403	No Clinical Signs, Symptoms or Conditions	A0701	Capturing Problem	26/07/2021	26/07/2021
									A072201	High Impedance	26/07/2021	26/07/2021
									A090805	Incorrect Measurement	26/07/2021	26/07/2021
							F1901	Additional Surgery	A0701	Capturing Problem	26/07/2021	26/07/2021
									A072201	High Impedance	26/07/2021	26/07/2021
									A090805	Incorrect Measurement	26/07/2021	26/07/2021
					CT2406	Implantable	E0717	Dyspnea	A0701	Capturing Problem	26/07/2021	26/07/2021
									A072201	High Impedance	26/07/2021	26/07/2021
									A090805	Incorrect Measurement	26/07/2021	26/07/2021
							E2403	No Clinical Signs, Symptoms or Conditions	A0701	Capturing Problem	26/07/2021	26/07/2021
									A072201	High Impedance	26/07/2021	26/07/2021
									A090805	Incorrect Measurement	26/07/2021	26/07/2021
							F1901	Additional Surgery	A0701	Capturing Problem	26/07/2021	26/07/2021
									A072201	High Impedance	26/07/2021	26/07/2021
									A090805	Incorrect Measurement	26/07/2021	26/07/2021
					CT2407	Active implantable	E0717	Dyspnea	A0701	Capturing Problem	26/07/2021	26/07/2021
									A072201	High Impedance	26/07/2021	26/07/2021
									A090805	Incorrect Measurement	26/07/2021	26/07/2021
							E2403	No Clinical Signs, Symptoms or Conditions	A0701	Capturing Problem	26/07/2021	26/07/2021
									A072201	High Impedance	26/07/2021	26/07/2021
									A090805	Incorrect Measurement	26/07/2021	26/07/2021
							F1901	Additional Surgery	A0701	Capturing Problem	26/07/2021	26/07/2021
									A072201	High Impedance	26/07/2021	26/07/2021
									A090805	Incorrect Measurement	26/07/2021	26/07/2021
					CT257	Implantable cardiac pacemakers	E0717	Dyspnea	A0701	Capturing Problem	26/07/2021	26/07/2021
									A072201	High Impedance	26/07/2021	26/07/2021
									A090805	Incorrect Measurement	26/07/2021	26/07/2021
							E2403	No Clinical Signs, Symptoms or Conditions	A0701	Capturing Problem	26/07/2021	26/07/2021
									A072201	High Impedance	26/07/2021	26/07/2021
									A090805	Incorrect Measurement	26/07/2021	26/07/2021
							F1901	Additional Surgery	A0701	Capturing Problem	26/07/2021	26/07/2021
									A072201	High Impedance	26/07/2021	26/07/2021
									A090805	Incorrect Measurement	26/07/2021	26/07/2021
					CT2727	Active implant devices/ancillaries	E0717	Dyspnea	A0701	Capturing Problem	26/07/2021	26/07/2021
									A072201	High Impedance	26/07/2021	26/07/2021
									A090805	Incorrect Measurement	26/07/2021	26/07/2021
							E2403	No Clinical Signs, Symptoms or Conditions	A0701	Capturing Problem	26/07/2021	26/07/2021
									A072201	High Impedance	26/07/2021	26/07/2021
									A090805	Incorrect Measurement	26/07/2021	26/07/2021
							F1901	Additional Surgery	A0701	Capturing Problem	26/07/2021	26/07/2021
									A072201	High Impedance	26/07/2021	26/07/2021
									A090805	Incorrect Measurement	26/07/2021	26/07/2021
					CT2750	Implantable cardiac electrical stimulators and associated devices	E0717	Dyspnea	A0701	Capturing Problem	26/07/2021	26/07/2021
									A072201	High Impedance	26/07/2021	26/07/2021
									A090805	Incorrect Measurement	26/07/2021	26/07/2021
							E2403	No Clinical Signs, Symptoms or Conditions	A0701	Capturing Problem	26/07/2021	26/07/2021

									A072201	High Impedance	26/07/2021	26/07/2021
									A090805	Incorrect Measurement	26/07/2021	26/07/2021
							F1901	Additional Surgery	A0701	Capturing Problem	26/07/2021	26/07/2021
									A072201	High Impedance	26/07/2021	26/07/2021
									A090805	Incorrect Measurement	26/07/2021	26/07/2021
					CT2866	Implantable cardiac electrical stimulators	E0717	Dyspnea	A0701	Capturing Problem	26/07/2021	26/07/2021
									A072201	High Impedance	26/07/2021	26/07/2021
									A090805	Incorrect Measurement	26/07/2021	26/07/2021
							E2403	No Clinical Signs, Symptoms or Conditions	A0701	Capturing Problem	26/07/2021	26/07/2021
									A072201	High Impedance	26/07/2021	26/07/2021
									A090805	Incorrect Measurement	26/07/2021	26/07/2021
							F1901	Additional Surgery	A0701	Capturing Problem	26/07/2021	26/07/2021
									A072201	High Impedance	26/07/2021	26/07/2021
									A090805	Incorrect Measurement	26/07/2021	26/07/2021
					CT2985	Active	E0717	Dyspnea	A0701	Capturing Problem	26/07/2021	26/07/2021
									A072201	High Impedance	26/07/2021	26/07/2021
									A090805	Incorrect Measurement	26/07/2021	26/07/2021
							E2403	No Clinical Signs, Symptoms or Conditions	A0701	Capturing Problem	26/07/2021	26/07/2021
									A072201	High Impedance	26/07/2021	26/07/2021
									A090805	Incorrect Measurement	26/07/2021	26/07/2021
							F1901	Additional Surgery	A0701	Capturing Problem	26/07/2021	26/07/2021
									A072201	High Impedance	26/07/2021	26/07/2021
									A090805	Incorrect Measurement	26/07/2021	26/07/2021
					CT3001	Implantable metal	E0717	Dyspnea	A0701	Capturing Problem	26/07/2021	26/07/2021
									A072201	High Impedance	26/07/2021	26/07/2021
									A090805	Incorrect Measurement	26/07/2021	26/07/2021
							E2403	No Clinical Signs, Symptoms or Conditions	A0701	Capturing Problem	26/07/2021	26/07/2021
									A072201	High Impedance	26/07/2021	26/07/2021
									A090805	Incorrect Measurement	26/07/2021	26/07/2021
							F1901	Additional Surgery	A0701	Capturing Problem	26/07/2021	26/07/2021
									A072201	High Impedance	26/07/2021	26/07/2021
									A090805	Incorrect Measurement	26/07/2021	26/07/2021
					CT3002	Implantable synthetic polymer	E0717	Dyspnea	A0701	Capturing Problem	26/07/2021	26/07/2021
									A072201	High Impedance	26/07/2021	26/07/2021
									A090805	Incorrect Measurement	26/07/2021	26/07/2021
							E2403	No Clinical Signs, Symptoms or Conditions	A0701	Capturing Problem	26/07/2021	26/07/2021
									A072201	High Impedance	26/07/2021	26/07/2021
									A090805	Incorrect Measurement	26/07/2021	26/07/2021
							F1901	Additional Surgery	A0701	Capturing Problem	26/07/2021	26/07/2021
									A072201	High Impedance	26/07/2021	26/07/2021
									A090805	Incorrect Measurement	26/07/2021	26/07/2021
					CT3006	Implantable inorganic material	E0717	Dyspnea	A0701	Capturing Problem	26/07/2021	26/07/2021
									A072201	High Impedance	26/07/2021	26/07/2021
									A090805	Incorrect Measurement	26/07/2021	26/07/2021
							E2403	No Clinical Signs, Symptoms or Conditions	A0701	Capturing Problem	26/07/2021	26/07/2021
									A072201	High Impedance	26/07/2021	26/07/2021
									A090805	Incorrect Measurement	26/07/2021	26/07/2021
							F1901	Additional Surgery	A0701	Capturing Problem	26/07/2021	26/07/2021
									A072201	High Impedance	26/07/2021	26/07/2021
									A090805	Incorrect Measurement	26/07/2021	26/07/2021
					CT314	Electrically powered	E0717	Dyspnea	A0701	Capturing Problem	26/07/2021	26/07/2021
									A072201	High Impedance	26/07/2021	26/07/2021
									A090805	Incorrect Measurement	26/07/2021	26/07/2021
							E2403	No Clinical Signs, Symptoms or Conditions	A0701	Capturing Problem	26/07/2021	26/07/2021
									A072201	High Impedance	26/07/2021	26/07/2021
									A090805	Incorrect Measurement	26/07/2021	26/07/2021
							F1901	Additional Surgery	A0701	Capturing Problem	26/07/2021	26/07/2021
									A072201	High Impedance	26/07/2021	26/07/2021
									A090805	Incorrect Measurement	26/07/2021	26/07/2021
					CT321	Long-term surgical invasive	E0717	Dyspnea	A0701	Capturing Problem	26/07/2021	26/07/2021
									A072201	High Impedance	26/07/2021	26/07/2021
									A090805	Incorrect Measurement	26/07/2021	26/07/2021
							E2403	No Clinical Signs, Symptoms or Conditions	A0701	Capturing Problem	26/07/2021	26/07/2021
									A072201	High Impedance	26/07/2021	26/07/2021
									A090805	Incorrect Measurement	26/07/2021	26/07/2021
							F1901	Additional Surgery	A0701	Capturing Problem	26/07/2021	26/07/2021
									A072201	High Impedance	26/07/2021	26/07/2021
									A090805	Incorrect Measurement	26/07/2021	26/07/2021
					CT449	Generators and associated devices	E0717	Dyspnea	A0701	Capturing Problem	26/07/2021	26/07/2021
									A072201	High Impedance	26/07/2021	26/07/2021
									A090805	Incorrect Measurement	26/07/2021	26/07/2021
							E2403	No Clinical Signs, Symptoms or Conditions	A0701	Capturing Problem	26/07/2021	26/07/2021
									A072201	High Impedance	26/07/2021	26/07/2021
									A090805	Incorrect Measurement	26/07/2021	26/07/2021
							F1901	Additional Surgery	A0701	Capturing Problem	26/07/2021	26/07/2021
									A072201	High Impedance	26/07/2021	26/07/2021
									A090805	Incorrect Measurement	26/07/2021	26/07/2021
					CT981	Single use	E0717	Dyspnea	A0701	Capturing Problem	26/07/2021	26/07/2021
									A072201	High Impedance	26/07/2021	26/07/2021
									A090805	Incorrect Measurement	26/07/2021	26/07/2021
							E2403	No Clinical Signs, Symptoms or Conditions	A0701	Capturing Problem	26/07/2021	26/07/2021
									A072201	High Impedance	26/07/2021	26/07/2021
									A090805	Incorrect Measurement	26/07/2021	26/07/2021
							F1901	Additional Surgery	A0701	Capturing Problem	26/07/2021	26/07/2021
									A072201	High Impedance	26/07/2021	26/07/2021
									A090805	Incorrect Measurement	26/07/2021	26/07/2021
					CT983	Surgical invasive	E0717	Dyspnea	A0701	Capturing Problem	26/07/2021	26/07/2021
									A072201	High Impedance	26/07/2021	26/07/2021

								A090805	Incorrect Measurement	26/07/2021	26/07/2021
						E2403	No Clinical Signs, Symptoms or Conditions	A0701	Capturing Problem	26/07/2021	26/07/2021
								A072201	High Impedance	26/07/2021	26/07/2021
								A090805	Incorrect Measurement	26/07/2021	26/07/2021
						F1901	Additional Surgery	A0701	Capturing Problem	26/07/2021	26/07/2021
								A072201	High Impedance	26/07/2021	26/07/2021
								A090805	Incorrect Measurement	26/07/2021	26/07/2021
6822	2022	All other reportable incidents	36241	Pacemaker/defibrillator lead	CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070909	Oversensing		
									A090801	Signal Artifact/Noise		
							F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A070909	Oversensing		
									A090801	Signal Artifact/Noise		
6823	2021	All other reportable incidents	35852	Single-chamber implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	12/10/2021	12/10/2021
							F1901	Additional Surgery	A070504	Premature Discharge of Battery	12/10/2021	12/10/2021
6824	2022	All other reportable incidents	47265	Dual-chamber implantable pacemaker, rate-responsive	CT141	Electrodes/Leads and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	01/05/2022	01/05/2022
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	01/05/2022	01/05/2022
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	01/05/2022	01/05/2022
					CT145	Cardiovascular devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	01/05/2022	01/05/2022
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	01/05/2022	01/05/2022
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	01/05/2022	01/05/2022
					CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	01/05/2022	01/05/2022
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	01/05/2022	01/05/2022
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	01/05/2022	01/05/2022
					CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	01/05/2022	01/05/2022
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	01/05/2022	01/05/2022
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	01/05/2022	01/05/2022
					CT233	Surgical	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	01/05/2022	01/05/2022
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	01/05/2022	01/05/2022
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	01/05/2022	01/05/2022
					CT2406	Implantable	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	01/05/2022	01/05/2022
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	01/05/2022	01/05/2022
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	01/05/2022	01/05/2022
					CT2407	Active implantable	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	01/05/2022	01/05/2022
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	01/05/2022	01/05/2022
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	01/05/2022	01/05/2022
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	01/05/2022	01/05/2022
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	01/05/2022	01/05/2022
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	01/05/2022	01/05/2022
					CT2727	Active implant devices/ancillaries	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	01/05/2022	01/05/2022
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	01/05/2022	01/05/2022
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	01/05/2022	01/05/2022
					CT2750	Implantable cardiac electrical stimulators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	01/05/2022	01/05/2022
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	01/05/2022	01/05/2022
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	01/05/2022	01/05/2022
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	01/05/2022	01/05/2022
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	01/05/2022	01/05/2022
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	01/05/2022	01/05/2022
					CT2985	Active	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	01/05/2022	01/05/2022
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	01/05/2022	01/05/2022
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	01/05/2022	01/05/2022
					CT3001	Implantable metal	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	01/05/2022	01/05/2022
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	01/05/2022	01/05/2022
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	01/05/2022	01/05/2022
					CT3002	Implantable synthetic polymer	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	01/05/2022	01/05/2022
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	01/05/2022	01/05/2022
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	01/05/2022	01/05/2022
					CT3006	Implantable inorganic material	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	01/05/2022	01/05/2022
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	01/05/2022	01/05/2022
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	01/05/2022	01/05/2022
					CT314	Electrically powered	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	01/05/2022	01/05/2022
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	01/05/2022	01/05/2022
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	01/05/2022	01/05/2022
					CT321	Long-term surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	01/05/2022	01/05/2022
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	01/05/2022	01/05/2022
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	01/05/2022	01/05/2022
					CT449	Generators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	01/05/2022	01/05/2022
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	01/05/2022	01/05/2022
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	01/05/2022	01/05/2022
					CT981	Single use	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	01/05/2022	01/05/2022
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	01/05/2022	01/05/2022
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	01/05/2022	01/05/2022
					CT983	Surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	01/05/2022	01/05/2022
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	01/05/2022	01/05/2022
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	01/05/2022	01/05/2022
6825	2022	All other reportable incidents	37265	Dual-chamber implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	11/07/2022	11/07/2022
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	11/07/2022	11/07/2022
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	11/07/2022	11/07/2022
6826	2022	All other reportable incidents	47265	Dual-chamber implantable pacemaker, rate-responsive	CT141	Electrodes/Leads and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT145	Cardiovascular devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT233	Surgical	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT2406	Implantable	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		

					CT2407	Active implantable	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT2727	Active implant devices/ancillaries	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT2750	Implantable cardiac electrical stimulators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT2985	Active	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT3001	Implantable metal	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT3002	Implantable synthetic polymer	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT3006	Implantable inorganic material	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT314	Electrically powered	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT321	Long-term surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT449	Generators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT981	Single use	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT983	Surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
6827	2021	All other reportable incidents	35223	Endocardial/interventricular septal pacing lead	CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070910	Undersensing	01/10/2021	01/10/2021
							F26	No Health Consequences or Impact	A070910	Undersensing	01/10/2021	01/10/2021
6828	2021	All other reportable incidents	47263	Cardiac resynchronization therapy implantable pacemaker	CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A0908	Incorrect, Inadequate or Imprecise Result or Readings	06/02/2022	06/02/2022
							F08	Hospitalization or Prolonged Hospitalization	A0908	Incorrect, Inadequate or Imprecise Result or Readings	06/02/2022	06/02/2022
							F1903	Device Explantation	A0908	Incorrect, Inadequate or Imprecise Result or Readings	06/02/2022	06/02/2022
					CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A0908	Incorrect, Inadequate or Imprecise Result or Readings	06/02/2022	06/02/2022
							F08	Hospitalization or Prolonged Hospitalization	A0908	Incorrect, Inadequate or Imprecise Result or Readings	06/02/2022	06/02/2022
							F1903	Device Explantation	A0908	Incorrect, Inadequate or Imprecise Result or Readings	06/02/2022	06/02/2022
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A0908	Incorrect, Inadequate or Imprecise Result or Readings	06/02/2022	06/02/2022
							F08	Hospitalization or Prolonged Hospitalization	A0908	Incorrect, Inadequate or Imprecise Result or Readings	06/02/2022	06/02/2022
							F1903	Device Explantation	A0908	Incorrect, Inadequate or Imprecise Result or Readings	06/02/2022	06/02/2022
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A0908	Incorrect, Inadequate or Imprecise Result or Readings	06/02/2022	06/02/2022
							F08	Hospitalization or Prolonged Hospitalization	A0908	Incorrect, Inadequate or Imprecise Result or Readings	06/02/2022	06/02/2022
							F1903	Device Explantation	A0908	Incorrect, Inadequate or Imprecise Result or Readings	06/02/2022	06/02/2022
6829	2022	All other reportable incidents	35852	Single-chamber implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A070101	Failure to Capture	26/05/2022	26/05/2022
									A070909	Oversensing	26/05/2022	26/05/2022
									A12	Connection Problem	26/05/2022	26/05/2022
							F1908	Prolonged Surgery	A070101	Failure to Capture	26/05/2022	26/05/2022
									A070909	Oversensing	26/05/2022	26/05/2022
									A12	Connection Problem	26/05/2022	26/05/2022
6830	2021	All other reportable incidents	47265	Dual-chamber implantable pacemaker, rate-responsive	CT141	Electrodes/Leads and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	10/02/2022	10/02/2022
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	10/02/2022	10/02/2022
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	10/02/2022	10/02/2022
					CT145	Cardiovascular devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	10/02/2022	10/02/2022
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	10/02/2022	10/02/2022
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	10/02/2022	10/02/2022
					CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	10/02/2022	10/02/2022
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	10/02/2022	10/02/2022
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	10/02/2022	10/02/2022
					CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	10/02/2022	10/02/2022
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	10/02/2022	10/02/2022
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	10/02/2022	10/02/2022
					CT233	Surgical	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	10/02/2022	10/02/2022
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	10/02/2022	10/02/2022
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	10/02/2022	10/02/2022
					CT2406	Implantable	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	10/02/2022	10/02/2022
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	10/02/2022	10/02/2022
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	10/02/2022	10/02/2022
					CT2407	Active implantable	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	10/02/2022	10/02/2022
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	10/02/2022	10/02/2022
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	10/02/2022	10/02/2022
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	10/02/2022	10/02/2022
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	10/02/2022	10/02/2022
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	10/02/2022	10/02/2022
					CT2727	Active implant devices/ancillaries	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	10/02/2022	10/02/2022
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	10/02/2022	10/02/2022
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	10/02/2022	10/02/2022
					CT2750	Implantable cardiac electrical stimulators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	10/02/2022	10/02/2022
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	10/02/2022	10/02/2022
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	10/02/2022	10/02/2022
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	10/02/2022	10/02/2022

							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	10/02/2022	10/02/2022
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	10/02/2022	10/02/2022
					CT2985	Active	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	10/02/2022	10/02/2022
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	10/02/2022	10/02/2022
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	10/02/2022	10/02/2022
					CT3001	Implantable metal	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	10/02/2022	10/02/2022
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	10/02/2022	10/02/2022
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	10/02/2022	10/02/2022
					CT3002	Implantable synthetic polymer	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	10/02/2022	10/02/2022
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	10/02/2022	10/02/2022
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	10/02/2022	10/02/2022
					CT3006	Implantable inorganic material	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	10/02/2022	10/02/2022
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	10/02/2022	10/02/2022
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	10/02/2022	10/02/2022
					CT314	Electrically powered	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	10/02/2022	10/02/2022
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	10/02/2022	10/02/2022
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	10/02/2022	10/02/2022
					CT321	Long-term surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	10/02/2022	10/02/2022
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	10/02/2022	10/02/2022
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	10/02/2022	10/02/2022
					CT449	Generators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	10/02/2022	10/02/2022
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	10/02/2022	10/02/2022
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	10/02/2022	10/02/2022
					CT981	Single use	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	10/02/2022	10/02/2022
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	10/02/2022	10/02/2022
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	10/02/2022	10/02/2022
					CT983	Surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	10/02/2022	10/02/2022
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	10/02/2022	10/02/2022
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	10/02/2022	10/02/2022
6831	2021	All other reportable incidents	36241	Pacemaker/defibrillator lead	CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A072201	High Impedance		
							F26	No Health Consequences or Impact	A072201	High Impedance		
6832	2021	All other reportable incidents	47265	Dual-chamber implantable pacemaker, rate-responsive	CT141	Electrodes/Leads and associated devices	E0112	Dizziness	A070504	Premature Discharge of Battery	23/09/2021	23/09/2021
							E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	23/09/2021	23/09/2021
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	23/09/2021	23/09/2021
							F1903	Device Explantation	A070504	Premature Discharge of Battery	23/09/2021	23/09/2021
					CT145	Cardiovascular devices	E0112	Dizziness	A070504	Premature Discharge of Battery	23/09/2021	23/09/2021
							E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	23/09/2021	23/09/2021
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	23/09/2021	23/09/2021
							F1903	Device Explantation	A070504	Premature Discharge of Battery	23/09/2021	23/09/2021
					CT1684	Cardiac pacemakers and associated devices	E0112	Dizziness	A070504	Premature Discharge of Battery	23/09/2021	23/09/2021
							E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	23/09/2021	23/09/2021
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	23/09/2021	23/09/2021
							F1903	Device Explantation	A070504	Premature Discharge of Battery	23/09/2021	23/09/2021
					CT1685	Cardiac pacemakers	E0112	Dizziness	A070504	Premature Discharge of Battery	23/09/2021	23/09/2021
							E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	23/09/2021	23/09/2021
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	23/09/2021	23/09/2021
							F1903	Device Explantation	A070504	Premature Discharge of Battery	23/09/2021	23/09/2021
					CT233	Surgical	E0112	Dizziness	A070504	Premature Discharge of Battery	23/09/2021	23/09/2021
							E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	23/09/2021	23/09/2021
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	23/09/2021	23/09/2021
							F1903	Device Explantation	A070504	Premature Discharge of Battery	23/09/2021	23/09/2021
					CT2406	Implantable	E0112	Dizziness	A070504	Premature Discharge of Battery	23/09/2021	23/09/2021
							E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	23/09/2021	23/09/2021
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	23/09/2021	23/09/2021
							F1903	Device Explantation	A070504	Premature Discharge of Battery	23/09/2021	23/09/2021
					CT2407	Active implantable	E0112	Dizziness	A070504	Premature Discharge of Battery	23/09/2021	23/09/2021
							E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	23/09/2021	23/09/2021
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	23/09/2021	23/09/2021
							F1903	Device Explantation	A070504	Premature Discharge of Battery	23/09/2021	23/09/2021
					CT257	Implantable cardiac pacemakers	E0112	Dizziness	A070504	Premature Discharge of Battery	23/09/2021	23/09/2021
							E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	23/09/2021	23/09/2021
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	23/09/2021	23/09/2021
							F1903	Device Explantation	A070504	Premature Discharge of Battery	23/09/2021	23/09/2021
					CT2727	Active implant devices/ancillaries	E0112	Dizziness	A070504	Premature Discharge of Battery	23/09/2021	23/09/2021
							E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	23/09/2021	23/09/2021
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	23/09/2021	23/09/2021
							F1903	Device Explantation	A070504	Premature Discharge of Battery	23/09/2021	23/09/2021
					CT2750	Implantable cardiac electrical stimulators and associated devices	E0112	Dizziness	A070504	Premature Discharge of Battery	23/09/2021	23/09/2021
							E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	23/09/2021	23/09/2021
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	23/09/2021	23/09/2021
							F1903	Device Explantation	A070504	Premature Discharge of Battery	23/09/2021	23/09/2021
					CT2866	Implantable cardiac electrical stimulators	E0112	Dizziness	A070504	Premature Discharge of Battery	23/09/2021	23/09/2021
							E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	23/09/2021	23/09/2021
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	23/09/2021	23/09/2021
							F1903	Device Explantation	A070504	Premature Discharge of Battery	23/09/2021	23/09/2021
					CT2985	Active	E0112	Dizziness	A070504	Premature Discharge of Battery	23/09/2021	23/09/2021
							E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	23/09/2021	23/09/2021
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	23/09/2021	23/09/2021
							F1903	Device Explantation	A070504	Premature Discharge of Battery	23/09/2021	23/09/2021
					CT3001	Implantable metal	E0112	Dizziness	A070504	Premature Discharge of Battery	23/09/2021	23/09/2021
							E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	23/09/2021	23/09/2021
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	23/09/2021	23/09/2021
							F1903	Device Explantation	A070504	Premature Discharge of Battery	23/09/2021	23/09/2021
					CT3002	Implantable synthetic polymer	E0112	Dizziness	A070504	Premature Discharge of Battery	23/09/2021	23/09/2021
							E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	23/09/2021	23/09/2021
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	23/09/2021	23/09/2021
							F1903	Device Explantation	A070504	Premature Discharge of Battery	23/09/2021	23/09/2021
					CT3006	Implantable inorganic material	E0112	Dizziness	A070504	Premature Discharge of Battery	23/09/2021	23/09/2021

							E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	23/09/2021	23/09/2021
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	23/09/2021	23/09/2021
							F1903	Device Explantation	A070504	Premature Discharge of Battery	23/09/2021	23/09/2021
					CT314	Electrically powered	E0112	Dizziness	A070504	Premature Discharge of Battery	23/09/2021	23/09/2021
							E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	23/09/2021	23/09/2021
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	23/09/2021	23/09/2021
							F1903	Device Explantation	A070504	Premature Discharge of Battery	23/09/2021	23/09/2021
					CT321	Long-term surgical invasive	E0112	Dizziness	A070504	Premature Discharge of Battery	23/09/2021	23/09/2021
							E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	23/09/2021	23/09/2021
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	23/09/2021	23/09/2021
							F1903	Device Explantation	A070504	Premature Discharge of Battery	23/09/2021	23/09/2021
					CT449	Generators and associated devices	E0112	Dizziness	A070504	Premature Discharge of Battery	23/09/2021	23/09/2021
							E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	23/09/2021	23/09/2021
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	23/09/2021	23/09/2021
							F1903	Device Explantation	A070504	Premature Discharge of Battery	23/09/2021	23/09/2021
					CT981	Single use	E0112	Dizziness	A070504	Premature Discharge of Battery	23/09/2021	23/09/2021
							E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	23/09/2021	23/09/2021
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	23/09/2021	23/09/2021
							F1903	Device Explantation	A070504	Premature Discharge of Battery	23/09/2021	23/09/2021
					CT983	Surgical invasive	E0112	Dizziness	A070504	Premature Discharge of Battery	23/09/2021	23/09/2021
							E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	23/09/2021	23/09/2021
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	23/09/2021	23/09/2021
							F1903	Device Explantation	A070504	Premature Discharge of Battery	23/09/2021	23/09/2021
6833	2022	All other reportable incidents	17882	Physiologic-monitoring defibrillation system	CT1684	Cardiac pacemakers and associated devices	E2401	Insufficient Information	A071204	No Pacing		
							E2403	No Clinical Signs, Symptoms or Conditions	A071204	No Pacing		
							F26	No Health Consequences or Impact	A071204	No Pacing		
6834	2022	All other reportable incidents	35039	Epicardial pacing lead	CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A051201	Device Dislodged or Dislocated		
									A070102	High Capture Threshold		
									A070104	Unstable Capture Threshold		
									A070901	Decreased Sensitivity		
									A070909	Oversensing		
									A072201	High Impedance		
							F08	Hospitalization or Prolonged Hospitalization	A051201	Device Dislodged or Dislocated		
									A070102	High Capture Threshold		
									A070104	Unstable Capture Threshold		
									A070901	Decreased Sensitivity		
									A070909	Oversensing		
									A072201	High Impedance		
							F12	Serious Injury/ Illness/ Impairment	A051201	Device Dislodged or Dislocated		
									A070102	High Capture Threshold		
									A070104	Unstable Capture Threshold		
									A070901	Decreased Sensitivity		
									A070909	Oversensing		
									A072201	High Impedance		
							F1905	Device Revision or Replacement	A051201	Device Dislodged or Dislocated		
									A070102	High Capture Threshold		
									A070104	Unstable Capture Threshold		
									A070901	Decreased Sensitivity		
									A070909	Oversensing		
									A072201	High Impedance		
6835	2022	All other reportable incidents	17882	Physiologic-monitoring defibrillation system	CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070803	Failure to Power Up		
							F27	Problem Identified During Non-Clinical Procedure	A070803	Failure to Power Up		
6836	2022	All other reportable incidents	17882	Physiologic-monitoring defibrillation system	CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A07	Electrical /Electronic Property Problem		
							F27	Problem Identified During Non-Clinical Procedure	A07	Electrical /Electronic Property Problem		
6837	2022	All other reportable incidents	35852	Single-chamber implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A070909	Oversensing	31/03/2022	31/03/2022
									A090801	Signal Artifact/Noise	31/03/2022	31/03/2022
							F13	Misdiagnosis/ Misclassification	A070909	Oversensing	31/03/2022	31/03/2022
									A090801	Signal Artifact/Noise	31/03/2022	31/03/2022
6838	2021	All other reportable incidents	47265	Dual-chamber implantable pacemaker, rate-responsive	CT141	Electrodes/Leads and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	26/01/2022	26/01/2022
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	26/01/2022	26/01/2022
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	26/01/2022	26/01/2022
					CT145	Cardiovascular devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	26/01/2022	26/01/2022
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	26/01/2022	26/01/2022
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	26/01/2022	26/01/2022
					CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	26/01/2022	26/01/2022
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	26/01/2022	26/01/2022
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	26/01/2022	26/01/2022
					CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	26/01/2022	26/01/2022
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	26/01/2022	26/01/2022
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	26/01/2022	26/01/2022
					CT233	Surgical	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	26/01/2022	26/01/2022
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	26/01/2022	26/01/2022
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	26/01/2022	26/01/2022
					CT2406	Implantable	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	26/01/2022	26/01/2022
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	26/01/2022	26/01/2022
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	26/01/2022	26/01/2022
					CT2407	Active implantable	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	26/01/2022	26/01/2022
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	26/01/2022	26/01/2022
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	26/01/2022	26/01/2022
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	26/01/2022	26/01/2022
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	26/01/2022	26/01/2022
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	26/01/2022	26/01/2022
					CT2727	Active implant devices/ancillaries	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	26/01/2022	26/01/2022
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	26/01/2022	26/01/2022
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	26/01/2022	26/01/2022
					CT2750	Implantable cardiac electrical stimulators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	26/01/2022	26/01/2022
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	26/01/2022	26/01/2022
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	26/01/2022	26/01/2022

					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	26/01/2022	26/01/2022
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	26/01/2022	26/01/2022
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	26/01/2022	26/01/2022
					CT2985	Active	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	26/01/2022	26/01/2022
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	26/01/2022	26/01/2022
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	26/01/2022	26/01/2022
					CT3001	Implantable metal	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	26/01/2022	26/01/2022
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	26/01/2022	26/01/2022
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	26/01/2022	26/01/2022
					CT3002	Implantable synthetic polymer	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	26/01/2022	26/01/2022
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	26/01/2022	26/01/2022
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	26/01/2022	26/01/2022
					CT3006	Implantable inorganic material	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	26/01/2022	26/01/2022
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	26/01/2022	26/01/2022
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	26/01/2022	26/01/2022
					CT314	Electrically powered	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	26/01/2022	26/01/2022
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	26/01/2022	26/01/2022
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	26/01/2022	26/01/2022
					CT321	Long-term surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	26/01/2022	26/01/2022
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	26/01/2022	26/01/2022
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	26/01/2022	26/01/2022
					CT449	Generators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	26/01/2022	26/01/2022
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	26/01/2022	26/01/2022
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	26/01/2022	26/01/2022
					CT981	Single use	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	26/01/2022	26/01/2022
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	26/01/2022	26/01/2022
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	26/01/2022	26/01/2022
					CT983	Surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	26/01/2022	26/01/2022
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	26/01/2022	26/01/2022
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	26/01/2022	26/01/2022
6839	2021	All other reportable incidents	47263	Cardiac resynchronization therapy implantable pacemaker	CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070102	High Capture Threshold		
									A070909	Oversensing		
									A090801	Signal Artifact/Noise		
							F26	No Health Consequences or Impact	A070102	High Capture Threshold		
									A070909	Oversensing		
									A090801	Signal Artifact/Noise		
					CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A070102	High Capture Threshold		
									A070909	Oversensing		
									A090801	Signal Artifact/Noise		
							F26	No Health Consequences or Impact	A070102	High Capture Threshold		
									A070909	Oversensing		
									A090801	Signal Artifact/Noise		
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A070102	High Capture Threshold		
									A070909	Oversensing		
									A090801	Signal Artifact/Noise		
							F26	No Health Consequences or Impact	A070102	High Capture Threshold		
									A070909	Oversensing		
									A090801	Signal Artifact/Noise		
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A070102	High Capture Threshold		
									A070909	Oversensing		
									A090801	Signal Artifact/Noise		
							F26	No Health Consequences or Impact	A070102	High Capture Threshold		
									A070909	Oversensing		
									A090801	Signal Artifact/Noise		
6840	2021	All other reportable incidents	35852	Single-chamber implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	12/06/2021	14/10/2021
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	12/06/2021	14/10/2021
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	12/06/2021	14/10/2021
6841	2021	All other reportable incidents	37265	Dual-chamber implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A072201	High Impedance		
							F26	No Health Consequences or Impact	A072201	High Impedance		
6842	2022	All other reportable incidents	47270	Cardiac resynchronization therapy implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
6843	2022	All other reportable incidents	35223	Endocardial/interventricular septal pacing lead	CT1684	Cardiac pacemakers and associated devices	E011903	Syncope/Fainting	A070101	Failure to Capture		
							E2403	No Clinical Signs, Symptoms or Conditions	A070101	Failure to Capture		
							F12	Serious Injury/ Illness/ Impairment	A070101	Failure to Capture		
6844	2021	All other reportable incidents	47265	Dual-chamber implantable pacemaker, rate-responsive	CT141	Electrodes/Leads and associated devices	E0623	Unspecified Heart Problem	A070909	Oversensing		
							E0717	Dyspnea	A070909	Oversensing		
							E2403	No Clinical Signs, Symptoms or Conditions	A070909	Oversensing		
							F23	Unexpected Medical Intervention	A070909	Oversensing		
					CT145	Cardiovascular devices	E0623	Unspecified Heart Problem	A070909	Oversensing		
							E0717	Dyspnea	A070909	Oversensing		
							E2403	No Clinical Signs, Symptoms or Conditions	A070909	Oversensing		
							F23	Unexpected Medical Intervention	A070909	Oversensing		
					CT1684	Cardiac pacemakers and associated devices	E0623	Unspecified Heart Problem	A070909	Oversensing		
							E0717	Dyspnea	A070909	Oversensing		
							E2403	No Clinical Signs, Symptoms or Conditions	A070909	Oversensing		
							F23	Unexpected Medical Intervention	A070909	Oversensing		
					CT1685	Cardiac pacemakers	E0623	Unspecified Heart Problem	A070909	Oversensing		
							E0717	Dyspnea	A070909	Oversensing		
							E2403	No Clinical Signs, Symptoms or Conditions	A070909	Oversensing		
							F23	Unexpected Medical Intervention	A070909	Oversensing		
					CT233	Surgical	E0623	Unspecified Heart Problem	A070909	Oversensing		
							E0717	Dyspnea	A070909	Oversensing		
							E2403	No Clinical Signs, Symptoms or Conditions	A070909	Oversensing		
							F23	Unexpected Medical Intervention	A070909	Oversensing		
					CT2406	Implantable	E0623	Unspecified Heart Problem	A070909	Oversensing		
							E0717	Dyspnea	A070909	Oversensing		
							E2403	No Clinical Signs, Symptoms or Conditions	A070909	Oversensing		
							F23	Unexpected Medical Intervention	A070909	Oversensing		

					CT2407	Active implantable	E0623	Unspecified Heart Problem	A070909	Oversensing		
							E0717	Dyspnea	A070909	Oversensing		
							E2403	No Clinical Signs, Symptoms or Conditions	A070909	Oversensing		
							F23	Unexpected Medical Intervention	A070909	Oversensing		
					CT257	Implantable cardiac pacemakers	E0623	Unspecified Heart Problem	A070909	Oversensing		
							E0717	Dyspnea	A070909	Oversensing		
							E2403	No Clinical Signs, Symptoms or Conditions	A070909	Oversensing		
							F23	Unexpected Medical Intervention	A070909	Oversensing		
					CT2727	Active implant devices/ancillaries	E0623	Unspecified Heart Problem	A070909	Oversensing		
							E0717	Dyspnea	A070909	Oversensing		
							E2403	No Clinical Signs, Symptoms or Conditions	A070909	Oversensing		
							F23	Unexpected Medical Intervention	A070909	Oversensing		
					CT2750	Implantable cardiac electrical stimulators and associated devices	E0623	Unspecified Heart Problem	A070909	Oversensing		
							E0717	Dyspnea	A070909	Oversensing		
							E2403	No Clinical Signs, Symptoms or Conditions	A070909	Oversensing		
							F23	Unexpected Medical Intervention	A070909	Oversensing		
					CT2866	Implantable cardiac electrical stimulators	E0623	Unspecified Heart Problem	A070909	Oversensing		
							E0717	Dyspnea	A070909	Oversensing		
							E2403	No Clinical Signs, Symptoms or Conditions	A070909	Oversensing		
							F23	Unexpected Medical Intervention	A070909	Oversensing		
					CT2985	Active	E0623	Unspecified Heart Problem	A070909	Oversensing		
							E0717	Dyspnea	A070909	Oversensing		
							E2403	No Clinical Signs, Symptoms or Conditions	A070909	Oversensing		
							F23	Unexpected Medical Intervention	A070909	Oversensing		
					CT3001	Implantable metal	E0623	Unspecified Heart Problem	A070909	Oversensing		
							E0717	Dyspnea	A070909	Oversensing		
							E2403	No Clinical Signs, Symptoms or Conditions	A070909	Oversensing		
							F23	Unexpected Medical Intervention	A070909	Oversensing		
					CT3002	Implantable synthetic polymer	E0623	Unspecified Heart Problem	A070909	Oversensing		
							E0717	Dyspnea	A070909	Oversensing		
							E2403	No Clinical Signs, Symptoms or Conditions	A070909	Oversensing		
							F23	Unexpected Medical Intervention	A070909	Oversensing		
					CT3006	Implantable inorganic material	E0623	Unspecified Heart Problem	A070909	Oversensing		
							E0717	Dyspnea	A070909	Oversensing		
							E2403	No Clinical Signs, Symptoms or Conditions	A070909	Oversensing		
							F23	Unexpected Medical Intervention	A070909	Oversensing		
					CT314	Electrically powered	E0623	Unspecified Heart Problem	A070909	Oversensing		
							E0717	Dyspnea	A070909	Oversensing		
							E2403	No Clinical Signs, Symptoms or Conditions	A070909	Oversensing		
							F23	Unexpected Medical Intervention	A070909	Oversensing		
					CT321	Long-term surgical invasive	E0623	Unspecified Heart Problem	A070909	Oversensing		
							E0717	Dyspnea	A070909	Oversensing		
							E2403	No Clinical Signs, Symptoms or Conditions	A070909	Oversensing		
							F23	Unexpected Medical Intervention	A070909	Oversensing		
					CT449	Generators and associated devices	E0623	Unspecified Heart Problem	A070909	Oversensing		
							E0717	Dyspnea	A070909	Oversensing		
							E2403	No Clinical Signs, Symptoms or Conditions	A070909	Oversensing		
							F23	Unexpected Medical Intervention	A070909	Oversensing		
					CT981	Single use	E0623	Unspecified Heart Problem	A070909	Oversensing		
							E0717	Dyspnea	A070909	Oversensing		
							E2403	No Clinical Signs, Symptoms or Conditions	A070909	Oversensing		
							F23	Unexpected Medical Intervention	A070909	Oversensing		
					CT983	Surgical invasive	E0623	Unspecified Heart Problem	A070909	Oversensing		
							E0717	Dyspnea	A070909	Oversensing		
							E2403	No Clinical Signs, Symptoms or Conditions	A070909	Oversensing		
							F23	Unexpected Medical Intervention	A070909	Oversensing		
6845	2021	All other reportable incidents	37265	Dual-chamber implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	13/10/2021	13/10/2021
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	13/10/2021	13/10/2021
							F1903	Device Explantation	A070504	Premature Discharge of Battery	13/10/2021	13/10/2021
6846	2021	All other reportable incidents	47265	Dual-chamber implantable pacemaker, rate-responsive	CT141	Electrodes/Leads and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	20/09/2021	20/09/2021
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	20/09/2021	20/09/2021
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	20/09/2021	20/09/2021
					CT145	Cardiovascular devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	20/09/2021	20/09/2021
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	20/09/2021	20/09/2021
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	20/09/2021	20/09/2021
					CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	20/09/2021	20/09/2021
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	20/09/2021	20/09/2021
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	20/09/2021	20/09/2021
					CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	20/09/2021	20/09/2021
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	20/09/2021	20/09/2021
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	20/09/2021	20/09/2021
					CT233	Surgical	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	20/09/2021	20/09/2021
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	20/09/2021	20/09/2021
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	20/09/2021	20/09/2021
					CT2406	Implantable	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	20/09/2021	20/09/2021
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	20/09/2021	20/09/2021
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	20/09/2021	20/09/2021
					CT2407	Active implantable	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	20/09/2021	20/09/2021
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	20/09/2021	20/09/2021
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	20/09/2021	20/09/2021
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	20/09/2021	20/09/2021
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	20/09/2021	20/09/2021
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	20/09/2021	20/09/2021
					CT2727	Active implant devices/ancillaries	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	20/09/2021	20/09/2021
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	20/09/2021	20/09/2021
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	20/09/2021	20/09/2021
					CT2750	Implantable cardiac electrical stimulators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	20/09/2021	20/09/2021
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	20/09/2021	20/09/2021

							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	20/09/2021	20/09/2021
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	20/09/2021	20/09/2021
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	20/09/2021	20/09/2021
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	20/09/2021	20/09/2021
					CT2985	Active	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	20/09/2021	20/09/2021
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	20/09/2021	20/09/2021
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	20/09/2021	20/09/2021
					CT3001	Implantable metal	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	20/09/2021	20/09/2021
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	20/09/2021	20/09/2021
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	20/09/2021	20/09/2021
					CT3002	Implantable synthetic polymer	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	20/09/2021	20/09/2021
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	20/09/2021	20/09/2021
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	20/09/2021	20/09/2021
					CT3006	Implantable inorganic material	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	20/09/2021	20/09/2021
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	20/09/2021	20/09/2021
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	20/09/2021	20/09/2021
					CT314	Electrically powered	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	20/09/2021	20/09/2021
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	20/09/2021	20/09/2021
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	20/09/2021	20/09/2021
					CT321	Long-term surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	20/09/2021	20/09/2021
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	20/09/2021	20/09/2021
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	20/09/2021	20/09/2021
					CT449	Generators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	20/09/2021	20/09/2021
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	20/09/2021	20/09/2021
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	20/09/2021	20/09/2021
					CT981	Single use	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	20/09/2021	20/09/2021
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	20/09/2021	20/09/2021
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	20/09/2021	20/09/2021
					CT983	Surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	20/09/2021	20/09/2021
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	20/09/2021	20/09/2021
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	20/09/2021	20/09/2021
6847	2021	All other reportable incidents	47265	Dual-chamber implantable pacemaker, rate-responsive	CT141	Electrodes/Leads and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	24/09/2021	24/09/2021
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	24/09/2021	24/09/2021
							F1903	Device Explantation	A070504	Premature Discharge of Battery	24/09/2021	24/09/2021
					CT145	Cardiovascular devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	24/09/2021	24/09/2021
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	24/09/2021	24/09/2021
							F1903	Device Explantation	A070504	Premature Discharge of Battery	24/09/2021	24/09/2021
					CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	24/09/2021	24/09/2021
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	24/09/2021	24/09/2021
							F1903	Device Explantation	A070504	Premature Discharge of Battery	24/09/2021	24/09/2021
					CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	24/09/2021	24/09/2021
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	24/09/2021	24/09/2021
							F1903	Device Explantation	A070504	Premature Discharge of Battery	24/09/2021	24/09/2021
					CT233	Surgical	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	24/09/2021	24/09/2021
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	24/09/2021	24/09/2021
							F1903	Device Explantation	A070504	Premature Discharge of Battery	24/09/2021	24/09/2021
					CT2406	Implantable	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	24/09/2021	24/09/2021
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	24/09/2021	24/09/2021
							F1903	Device Explantation	A070504	Premature Discharge of Battery	24/09/2021	24/09/2021
					CT2407	Active implantable	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	24/09/2021	24/09/2021
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	24/09/2021	24/09/2021
							F1903	Device Explantation	A070504	Premature Discharge of Battery	24/09/2021	24/09/2021
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	24/09/2021	24/09/2021
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	24/09/2021	24/09/2021
							F1903	Device Explantation	A070504	Premature Discharge of Battery	24/09/2021	24/09/2021
					CT2727	Active implant devices/ancillaries	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	24/09/2021	24/09/2021
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	24/09/2021	24/09/2021
							F1903	Device Explantation	A070504	Premature Discharge of Battery	24/09/2021	24/09/2021
					CT2750	Implantable cardiac electrical stimulators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	24/09/2021	24/09/2021
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	24/09/2021	24/09/2021
							F1903	Device Explantation	A070504	Premature Discharge of Battery	24/09/2021	24/09/2021
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	24/09/2021	24/09/2021
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	24/09/2021	24/09/2021
							F1903	Device Explantation	A070504	Premature Discharge of Battery	24/09/2021	24/09/2021
					CT2985	Active	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	24/09/2021	24/09/2021
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	24/09/2021	24/09/2021
							F1903	Device Explantation	A070504	Premature Discharge of Battery	24/09/2021	24/09/2021
					CT3001	Implantable metal	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	24/09/2021	24/09/2021
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	24/09/2021	24/09/2021
							F1903	Device Explantation	A070504	Premature Discharge of Battery	24/09/2021	24/09/2021
					CT3002	Implantable synthetic polymer	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	24/09/2021	24/09/2021
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	24/09/2021	24/09/2021
							F1903	Device Explantation	A070504	Premature Discharge of Battery	24/09/2021	24/09/2021
					CT3006	Implantable inorganic material	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	24/09/2021	24/09/2021
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	24/09/2021	24/09/2021
							F1903	Device Explantation	A070504	Premature Discharge of Battery	24/09/2021	24/09/2021
					CT314	Electrically powered	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	24/09/2021	24/09/2021
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	24/09/2021	24/09/2021
							F1903	Device Explantation	A070504	Premature Discharge of Battery	24/09/2021	24/09/2021
					CT321	Long-term surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	24/09/2021	24/09/2021
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	24/09/2021	24/09/2021
							F1903	Device Explantation	A070504	Premature Discharge of Battery	24/09/2021	24/09/2021
					CT449	Generators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	24/09/2021	24/09/2021
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	24/09/2021	24/09/2021
							F1903	Device Explantation	A070504	Premature Discharge of Battery	24/09/2021	24/09/2021
					CT981	Single use	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	24/09/2021	24/09/2021
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	24/09/2021	24/09/2021
							F1903	Device Explantation	A070504	Premature Discharge of Battery	24/09/2021	24/09/2021

					CT983	Surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	24/09/2021	24/09/2021
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	24/09/2021	24/09/2021
							F1903	Device Explantation	A070504	Premature Discharge of Battery	24/09/2021	24/09/2021
6848	2021	All other reportable incidents	47267	Single-chamber implantable pacemaker, rate-responsive	CT141	Electrodes/Leads and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070909	Oversensing		
									A072201	High Impedance		
							F11	Non-Serious Injury/ Illness/ Impairment	A070909	Oversensing		
									A072201	High Impedance		
					CT145	Cardiovascular devices	E2403	No Clinical Signs, Symptoms or Conditions	A070909	Oversensing		
									A072201	High Impedance		
							F11	Non-Serious Injury/ Illness/ Impairment	A070909	Oversensing		
									A072201	High Impedance		
					CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070909	Oversensing		
									A072201	High Impedance		
							F11	Non-Serious Injury/ Illness/ Impairment	A070909	Oversensing		
									A072201	High Impedance		
					CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A070909	Oversensing		
									A072201	High Impedance		
							F11	Non-Serious Injury/ Illness/ Impairment	A070909	Oversensing		
									A072201	High Impedance		
					CT233	Surgical	E2403	No Clinical Signs, Symptoms or Conditions	A070909	Oversensing		
									A072201	High Impedance		
							F11	Non-Serious Injury/ Illness/ Impairment	A070909	Oversensing		
									A072201	High Impedance		
					CT2406	Implantable	E2403	No Clinical Signs, Symptoms or Conditions	A070909	Oversensing		
									A072201	High Impedance		
							F11	Non-Serious Injury/ Illness/ Impairment	A070909	Oversensing		
									A072201	High Impedance		
					CT2407	Active implantable	E2403	No Clinical Signs, Symptoms or Conditions	A070909	Oversensing		
									A072201	High Impedance		
							F11	Non-Serious Injury/ Illness/ Impairment	A070909	Oversensing		
									A072201	High Impedance		
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A070909	Oversensing		
									A072201	High Impedance		
							F11	Non-Serious Injury/ Illness/ Impairment	A070909	Oversensing		
									A072201	High Impedance		
					CT2727	Active implant devices/ancillaries	E2403	No Clinical Signs, Symptoms or Conditions	A070909	Oversensing		
									A072201	High Impedance		
							F11	Non-Serious Injury/ Illness/ Impairment	A070909	Oversensing		
									A072201	High Impedance		
					CT2750	Implantable cardiac electrical stimulators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070909	Oversensing		
									A072201	High Impedance		
							F11	Non-Serious Injury/ Illness/ Impairment	A070909	Oversensing		
									A072201	High Impedance		
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A070909	Oversensing		
									A072201	High Impedance		
							F11	Non-Serious Injury/ Illness/ Impairment	A070909	Oversensing		
									A072201	High Impedance		
					CT2985	Active	E2403	No Clinical Signs, Symptoms or Conditions	A070909	Oversensing		
									A072201	High Impedance		
							F11	Non-Serious Injury/ Illness/ Impairment	A070909	Oversensing		
									A072201	High Impedance		
					CT3001	Implantable metal	E2403	No Clinical Signs, Symptoms or Conditions	A070909	Oversensing		
									A072201	High Impedance		
							F11	Non-Serious Injury/ Illness/ Impairment	A070909	Oversensing		
									A072201	High Impedance		
					CT3002	Implantable synthetic polymer	E2403	No Clinical Signs, Symptoms or Conditions	A070909	Oversensing		
									A072201	High Impedance		
							F11	Non-Serious Injury/ Illness/ Impairment	A070909	Oversensing		
									A072201	High Impedance		
					CT3006	Implantable inorganic material	E2403	No Clinical Signs, Symptoms or Conditions	A070909	Oversensing		
									A072201	High Impedance		
							F11	Non-Serious Injury/ Illness/ Impairment	A070909	Oversensing		
									A072201	High Impedance		
					CT314	Electrically powered	E2403	No Clinical Signs, Symptoms or Conditions	A070909	Oversensing		
									A072201	High Impedance		
							F11	Non-Serious Injury/ Illness/ Impairment	A070909	Oversensing		
									A072201	High Impedance		
					CT321	Long-term surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A070909	Oversensing		
									A072201	High Impedance		
							F11	Non-Serious Injury/ Illness/ Impairment	A070909	Oversensing		
									A072201	High Impedance		
					CT449	Generators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070909	Oversensing		
									A072201	High Impedance		
							F11	Non-Serious Injury/ Illness/ Impairment	A070909	Oversensing		
									A072201	High Impedance		
					CT981	Single use	E2403	No Clinical Signs, Symptoms or Conditions	A070909	Oversensing		
									A072201	High Impedance		
							F11	Non-Serious Injury/ Illness/ Impairment	A070909	Oversensing		
									A072201	High Impedance		
					CT983	Surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A070909	Oversensing		
									A072201	High Impedance		
							F11	Non-Serious Injury/ Illness/ Impairment	A070909	Oversensing		
									A072201	High Impedance		
6849	2021	All other reportable incidents	35852	Single-chamber implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	24/12/2021	24/12/2021
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	24/12/2021	24/12/2021
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	24/12/2021	24/12/2021
6850	2021	All other reportable incidents	47265	Dual-chamber implantable pacemaker, rate-responsive	CT141	Electrodes/Leads and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A0908	Incorrect, Inadequate or Imprecise Result or Readings	07/10/2021	07/10/2021

							F08	Hospitalization or Prolonged Hospitalization	A0908	Incorrect, Inadequate or Imprecise Result or Readings	07/10/2021	07/10/2021
							F1905	Device Revision or Replacement	A0908	Incorrect, Inadequate or Imprecise Result or Readings	07/10/2021	07/10/2021
					CT145	Cardiovascular devices	E2403	No Clinical Signs, Symptoms or Conditions	A0908	Incorrect, Inadequate or Imprecise Result or Readings	07/10/2021	07/10/2021
							F08	Hospitalization or Prolonged Hospitalization	A0908	Incorrect, Inadequate or Imprecise Result or Readings	07/10/2021	07/10/2021
							F1905	Device Revision or Replacement	A0908	Incorrect, Inadequate or Imprecise Result or Readings	07/10/2021	07/10/2021
					CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A0908	Incorrect, Inadequate or Imprecise Result or Readings	07/10/2021	07/10/2021
							F08	Hospitalization or Prolonged Hospitalization	A0908	Incorrect, Inadequate or Imprecise Result or Readings	07/10/2021	07/10/2021
							F1905	Device Revision or Replacement	A0908	Incorrect, Inadequate or Imprecise Result or Readings	07/10/2021	07/10/2021
					CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A0908	Incorrect, Inadequate or Imprecise Result or Readings	07/10/2021	07/10/2021
							F08	Hospitalization or Prolonged Hospitalization	A0908	Incorrect, Inadequate or Imprecise Result or Readings	07/10/2021	07/10/2021
							F1905	Device Revision or Replacement	A0908	Incorrect, Inadequate or Imprecise Result or Readings	07/10/2021	07/10/2021
					CT233	Surgical	E2403	No Clinical Signs, Symptoms or Conditions	A0908	Incorrect, Inadequate or Imprecise Result or Readings	07/10/2021	07/10/2021
							F08	Hospitalization or Prolonged Hospitalization	A0908	Incorrect, Inadequate or Imprecise Result or Readings	07/10/2021	07/10/2021
							F1905	Device Revision or Replacement	A0908	Incorrect, Inadequate or Imprecise Result or Readings	07/10/2021	07/10/2021
					CT2406	Implantable	E2403	No Clinical Signs, Symptoms or Conditions	A0908	Incorrect, Inadequate or Imprecise Result or Readings	07/10/2021	07/10/2021
							F08	Hospitalization or Prolonged Hospitalization	A0908	Incorrect, Inadequate or Imprecise Result or Readings	07/10/2021	07/10/2021
							F1905	Device Revision or Replacement	A0908	Incorrect, Inadequate or Imprecise Result or Readings	07/10/2021	07/10/2021
					CT2407	Active implantable	E2403	No Clinical Signs, Symptoms or Conditions	A0908	Incorrect, Inadequate or Imprecise Result or Readings	07/10/2021	07/10/2021
							F08	Hospitalization or Prolonged Hospitalization	A0908	Incorrect, Inadequate or Imprecise Result or Readings	07/10/2021	07/10/2021
							F1905	Device Revision or Replacement	A0908	Incorrect, Inadequate or Imprecise Result or Readings	07/10/2021	07/10/2021
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A0908	Incorrect, Inadequate or Imprecise Result or Readings	07/10/2021	07/10/2021
							F08	Hospitalization or Prolonged Hospitalization	A0908	Incorrect, Inadequate or Imprecise Result or Readings	07/10/2021	07/10/2021
							F1905	Device Revision or Replacement	A0908	Incorrect, Inadequate or Imprecise Result or Readings	07/10/2021	07/10/2021
					CT2727	Active implant devices/ancillaries	E2403	No Clinical Signs, Symptoms or Conditions	A0908	Incorrect, Inadequate or Imprecise Result or Readings	07/10/2021	07/10/2021
							F08	Hospitalization or Prolonged Hospitalization	A0908	Incorrect, Inadequate or Imprecise Result or Readings	07/10/2021	07/10/2021
							F1905	Device Revision or Replacement	A0908	Incorrect, Inadequate or Imprecise Result or Readings	07/10/2021	07/10/2021
					CT2750	Implantable cardiac electrical stimulators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A0908	Incorrect, Inadequate or Imprecise Result or Readings	07/10/2021	07/10/2021
							F08	Hospitalization or Prolonged Hospitalization	A0908	Incorrect, Inadequate or Imprecise Result or Readings	07/10/2021	07/10/2021
							F1905	Device Revision or Replacement	A0908	Incorrect, Inadequate or Imprecise Result or Readings	07/10/2021	07/10/2021
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A0908	Incorrect, Inadequate or Imprecise Result or Readings	07/10/2021	07/10/2021
							F08	Hospitalization or Prolonged Hospitalization	A0908	Incorrect, Inadequate or Imprecise Result or Readings	07/10/2021	07/10/2021
							F1905	Device Revision or Replacement	A0908	Incorrect, Inadequate or Imprecise Result or Readings	07/10/2021	07/10/2021
					CT2985	Active	E2403	No Clinical Signs, Symptoms or Conditions	A0908	Incorrect, Inadequate or Imprecise Result or Readings	07/10/2021	07/10/2021
							F08	Hospitalization or Prolonged Hospitalization	A0908	Incorrect, Inadequate or Imprecise Result or Readings	07/10/2021	07/10/2021
							F1905	Device Revision or Replacement	A0908	Incorrect, Inadequate or Imprecise Result or Readings	07/10/2021	07/10/2021
					CT3001	Implantable metal	E2403	No Clinical Signs, Symptoms or Conditions	A0908	Incorrect, Inadequate or Imprecise Result or Readings	07/10/2021	07/10/2021
							F08	Hospitalization or Prolonged Hospitalization	A0908	Incorrect, Inadequate or Imprecise Result or Readings	07/10/2021	07/10/2021
							F1905	Device Revision or Replacement	A0908	Incorrect, Inadequate or Imprecise Result or Readings	07/10/2021	07/10/2021
					CT3002	Implantable synthetic polymer	E2403	No Clinical Signs, Symptoms or Conditions	A0908	Incorrect, Inadequate or Imprecise Result or Readings	07/10/2021	07/10/2021
							F08	Hospitalization or Prolonged Hospitalization	A0908	Incorrect, Inadequate or Imprecise Result or Readings	07/10/2021	07/10/2021
							F1905	Device Revision or Replacement	A0908	Incorrect, Inadequate or Imprecise Result or Readings	07/10/2021	07/10/2021
					CT3006	Implantable inorganic material	E2403	No Clinical Signs, Symptoms or Conditions	A0908	Incorrect, Inadequate or Imprecise Result or Readings	07/10/2021	07/10/2021
							F08	Hospitalization or Prolonged Hospitalization	A0908	Incorrect, Inadequate or Imprecise Result or Readings	07/10/2021	07/10/2021
							F1905	Device Revision or Replacement	A0908	Incorrect, Inadequate or Imprecise Result or Readings	07/10/2021	07/10/2021
					CT314	Electrically powered	E2403	No Clinical Signs, Symptoms or Conditions	A0908	Incorrect, Inadequate or Imprecise Result or Readings	07/10/2021	07/10/2021
							F08	Hospitalization or Prolonged Hospitalization	A0908	Incorrect, Inadequate or Imprecise Result or Readings	07/10/2021	07/10/2021
							F1905	Device Revision or Replacement	A0908	Incorrect, Inadequate or Imprecise Result or Readings	07/10/2021	07/10/2021
					CT321	Long-term surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A0908	Incorrect, Inadequate or Imprecise Result or Readings	07/10/2021	07/10/2021
							F08	Hospitalization or Prolonged Hospitalization	A0908	Incorrect, Inadequate or Imprecise Result or Readings	07/10/2021	07/10/2021
							F1905	Device Revision or Replacement	A0908	Incorrect, Inadequate or Imprecise Result or Readings	07/10/2021	07/10/2021

					CT449	Generators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A0908	Incorrect, Inadequate or Imprecise Result or Readings	07/10/2021	07/10/2021
							F08	Hospitalization or Prolonged Hospitalization	A0908	Incorrect, Inadequate or Imprecise Result or Readings	07/10/2021	07/10/2021
							F1905	Device Revision or Replacement	A0908	Incorrect, Inadequate or Imprecise Result or Readings	07/10/2021	07/10/2021
					CT981	Single use	E2403	No Clinical Signs, Symptoms or Conditions	A0908	Incorrect, Inadequate or Imprecise Result or Readings	07/10/2021	07/10/2021
							F08	Hospitalization or Prolonged Hospitalization	A0908	Incorrect, Inadequate or Imprecise Result or Readings	07/10/2021	07/10/2021
							F1905	Device Revision or Replacement	A0908	Incorrect, Inadequate or Imprecise Result or Readings	07/10/2021	07/10/2021
					CT983	Surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A0908	Incorrect, Inadequate or Imprecise Result or Readings	07/10/2021	07/10/2021
							F08	Hospitalization or Prolonged Hospitalization	A0908	Incorrect, Inadequate or Imprecise Result or Readings	07/10/2021	07/10/2021
							F1905	Device Revision or Replacement	A0908	Incorrect, Inadequate or Imprecise Result or Readings	07/10/2021	07/10/2021
6851	2021	All other reportable incidents	37265	Dual-chamber implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E2104	Electric Shock	A0712	Pacing Problem		
									A071302	Inappropriate/Inadequate Shock/Stimulation		
							E2403	No Clinical Signs, Symptoms or Conditions	A0712	Pacing Problem		
									A071302	Inappropriate/Inadequate Shock/Stimulation		
							F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A0712	Pacing Problem		
									A071302	Inappropriate/Inadequate Shock/Stimulation		
6852	2021	All other reportable incidents	47265	Dual-chamber implantable pacemaker, rate-responsive	CT141	Electrodes/Leads and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	26/11/2021	26/11/2021
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	26/11/2021	26/11/2021
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	26/11/2021	26/11/2021
					CT145	Cardiovascular devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	26/11/2021	26/11/2021
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	26/11/2021	26/11/2021
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	26/11/2021	26/11/2021
					CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	26/11/2021	26/11/2021
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	26/11/2021	26/11/2021
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	26/11/2021	26/11/2021
					CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	26/11/2021	26/11/2021
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	26/11/2021	26/11/2021
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	26/11/2021	26/11/2021
					CT233	Surgical	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	26/11/2021	26/11/2021
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	26/11/2021	26/11/2021
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	26/11/2021	26/11/2021
					CT2406	Implantable	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	26/11/2021	26/11/2021
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	26/11/2021	26/11/2021
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	26/11/2021	26/11/2021
					CT2407	Active implantable	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	26/11/2021	26/11/2021
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	26/11/2021	26/11/2021
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	26/11/2021	26/11/2021
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	26/11/2021	26/11/2021
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	26/11/2021	26/11/2021
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	26/11/2021	26/11/2021
					CT2727	Active implant devices/ancillaries	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	26/11/2021	26/11/2021
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	26/11/2021	26/11/2021
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	26/11/2021	26/11/2021
					CT2750	Implantable cardiac electrical stimulators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	26/11/2021	26/11/2021
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	26/11/2021	26/11/2021
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	26/11/2021	26/11/2021
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	26/11/2021	26/11/2021
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	26/11/2021	26/11/2021
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	26/11/2021	26/11/2021
					CT2985	Active	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	26/11/2021	26/11/2021
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	26/11/2021	26/11/2021
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	26/11/2021	26/11/2021
					CT3001	Implantable metal	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	26/11/2021	26/11/2021
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	26/11/2021	26/11/2021
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	26/11/2021	26/11/2021
					CT3002	Implantable synthetic polymer	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	26/11/2021	26/11/2021
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	26/11/2021	26/11/2021
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	26/11/2021	26/11/2021
					CT3006	Implantable inorganic material	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	26/11/2021	26/11/2021
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	26/11/2021	26/11/2021
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	26/11/2021	26/11/2021
					CT314	Electrically powered	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	26/11/2021	26/11/2021
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	26/11/2021	26/11/2021
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	26/11/2021	26/11/2021
					CT321	Long-term surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	26/11/2021	26/11/2021
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	26/11/2021	26/11/2021
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	26/11/2021	26/11/2021
					CT449	Generators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	26/11/2021	26/11/2021
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	26/11/2021	26/11/2021
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	26/11/2021	26/11/2021
					CT981	Single use	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	26/11/2021	26/11/2021
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	26/11/2021	26/11/2021
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	26/11/2021	26/11/2021
					CT983	Surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	26/11/2021	26/11/2021
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	26/11/2021	26/11/2021
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	26/11/2021	26/11/2021
6853	2021	All other reportable incidents	47265	Dual-chamber implantable pacemaker, rate-responsive	CT141	Electrodes/Leads and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem	05/11/2021	05/11/2021
									A070504	Premature Discharge of Battery	05/11/2021	05/11/2021
							F08	Hospitalization or Prolonged Hospitalization	A0705	Battery Problem	05/11/2021	05/11/2021
									A070504	Premature Discharge of Battery	05/11/2021	05/11/2021
							F1905	Device Revision or Replacement	A0705	Battery Problem	05/11/2021	05/11/2021
									A070504	Premature Discharge of Battery	05/11/2021	05/11/2021

CT145	Cardiovascular devices	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem	05/11/2021	05/11/2021
				A070504	Premature Discharge of Battery	05/11/2021	05/11/2021
		F08	Hospitalization or Prolonged Hospitalization	A0705	Battery Problem	05/11/2021	05/11/2021
				A070504	Premature Discharge of Battery	05/11/2021	05/11/2021
		F1905	Device Revision or Replacement	A0705	Battery Problem	05/11/2021	05/11/2021
				A070504	Premature Discharge of Battery	05/11/2021	05/11/2021
CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem	05/11/2021	05/11/2021
				A070504	Premature Discharge of Battery	05/11/2021	05/11/2021
		F08	Hospitalization or Prolonged Hospitalization	A0705	Battery Problem	05/11/2021	05/11/2021
				A070504	Premature Discharge of Battery	05/11/2021	05/11/2021
		F1905	Device Revision or Replacement	A0705	Battery Problem	05/11/2021	05/11/2021
				A070504	Premature Discharge of Battery	05/11/2021	05/11/2021
CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem	05/11/2021	05/11/2021
				A070504	Premature Discharge of Battery	05/11/2021	05/11/2021
		F08	Hospitalization or Prolonged Hospitalization	A0705	Battery Problem	05/11/2021	05/11/2021
				A070504	Premature Discharge of Battery	05/11/2021	05/11/2021
		F1905	Device Revision or Replacement	A0705	Battery Problem	05/11/2021	05/11/2021
				A070504	Premature Discharge of Battery	05/11/2021	05/11/2021
CT233	Surgical	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem	05/11/2021	05/11/2021
				A070504	Premature Discharge of Battery	05/11/2021	05/11/2021
		F08	Hospitalization or Prolonged Hospitalization	A0705	Battery Problem	05/11/2021	05/11/2021
				A070504	Premature Discharge of Battery	05/11/2021	05/11/2021
		F1905	Device Revision or Replacement	A0705	Battery Problem	05/11/2021	05/11/2021
				A070504	Premature Discharge of Battery	05/11/2021	05/11/2021
CT2406	Implantable	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem	05/11/2021	05/11/2021
				A070504	Premature Discharge of Battery	05/11/2021	05/11/2021
		F08	Hospitalization or Prolonged Hospitalization	A0705	Battery Problem	05/11/2021	05/11/2021
				A070504	Premature Discharge of Battery	05/11/2021	05/11/2021
		F1905	Device Revision or Replacement	A0705	Battery Problem	05/11/2021	05/11/2021
				A070504	Premature Discharge of Battery	05/11/2021	05/11/2021
CT2407	Active implantable	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem	05/11/2021	05/11/2021
				A070504	Premature Discharge of Battery	05/11/2021	05/11/2021
		F08	Hospitalization or Prolonged Hospitalization	A0705	Battery Problem	05/11/2021	05/11/2021
				A070504	Premature Discharge of Battery	05/11/2021	05/11/2021
		F1905	Device Revision or Replacement	A0705	Battery Problem	05/11/2021	05/11/2021
				A070504	Premature Discharge of Battery	05/11/2021	05/11/2021
CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem	05/11/2021	05/11/2021
				A070504	Premature Discharge of Battery	05/11/2021	05/11/2021
		F08	Hospitalization or Prolonged Hospitalization	A0705	Battery Problem	05/11/2021	05/11/2021
				A070504	Premature Discharge of Battery	05/11/2021	05/11/2021
		F1905	Device Revision or Replacement	A0705	Battery Problem	05/11/2021	05/11/2021
				A070504	Premature Discharge of Battery	05/11/2021	05/11/2021
CT2727	Active implant devices/ancillaries	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem	05/11/2021	05/11/2021
				A070504	Premature Discharge of Battery	05/11/2021	05/11/2021
		F08	Hospitalization or Prolonged Hospitalization	A0705	Battery Problem	05/11/2021	05/11/2021
				A070504	Premature Discharge of Battery	05/11/2021	05/11/2021
		F1905	Device Revision or Replacement	A0705	Battery Problem	05/11/2021	05/11/2021
				A070504	Premature Discharge of Battery	05/11/2021	05/11/2021
CT2750	Implantable cardiac electrical stimulators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem	05/11/2021	05/11/2021
				A070504	Premature Discharge of Battery	05/11/2021	05/11/2021
		F08	Hospitalization or Prolonged Hospitalization	A0705	Battery Problem	05/11/2021	05/11/2021
				A070504	Premature Discharge of Battery	05/11/2021	05/11/2021
		F1905	Device Revision or Replacement	A0705	Battery Problem	05/11/2021	05/11/2021
				A070504	Premature Discharge of Battery	05/11/2021	05/11/2021
CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem	05/11/2021	05/11/2021
				A070504	Premature Discharge of Battery	05/11/2021	05/11/2021
		F08	Hospitalization or Prolonged Hospitalization	A0705	Battery Problem	05/11/2021	05/11/2021
				A070504	Premature Discharge of Battery	05/11/2021	05/11/2021
		F1905	Device Revision or Replacement	A0705	Battery Problem	05/11/2021	05/11/2021
				A070504	Premature Discharge of Battery	05/11/2021	05/11/2021
CT2985	Active	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem	05/11/2021	05/11/2021
				A070504	Premature Discharge of Battery	05/11/2021	05/11/2021
		F08	Hospitalization or Prolonged Hospitalization	A0705	Battery Problem	05/11/2021	05/11/2021
				A070504	Premature Discharge of Battery	05/11/2021	05/11/2021
		F1905	Device Revision or Replacement	A0705	Battery Problem	05/11/2021	05/11/2021
				A070504	Premature Discharge of Battery	05/11/2021	05/11/2021
CT3001	Implantable metal	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem	05/11/2021	05/11/2021
				A070504	Premature Discharge of Battery	05/11/2021	05/11/2021
		F08	Hospitalization or Prolonged Hospitalization	A0705	Battery Problem	05/11/2021	05/11/2021
				A070504	Premature Discharge of Battery	05/11/2021	05/11/2021
		F1905	Device Revision or Replacement	A0705	Battery Problem	05/11/2021	05/11/2021
				A070504	Premature Discharge of Battery	05/11/2021	05/11/2021
CT3002	Implantable synthetic polymer	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem	05/11/2021	05/11/2021
				A070504	Premature Discharge of Battery	05/11/2021	05/11/2021
		F08	Hospitalization or Prolonged Hospitalization	A0705	Battery Problem	05/11/2021	05/11/2021
				A070504	Premature Discharge of Battery	05/11/2021	05/11/2021
		F1905	Device Revision or Replacement	A0705	Battery Problem	05/11/2021	05/11/2021
				A070504	Premature Discharge of Battery	05/11/2021	05/11/2021
CT3006	Implantable inorganic material	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem	05/11/2021	05/11/2021
				A070504	Premature Discharge of Battery	05/11/2021	05/11/2021
		F08	Hospitalization or Prolonged Hospitalization	A0705	Battery Problem	05/11/2021	05/11/2021
				A070504	Premature Discharge of Battery	05/11/2021	05/11/2021
		F1905	Device Revision or Replacement	A0705	Battery Problem	05/11/2021	05/11/2021
				A070504	Premature Discharge of Battery	05/11/2021	05/11/2021
CT314	Electrically powered	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem	05/11/2021	05/11/2021
				A070504	Premature Discharge of Battery	05/11/2021	05/11/2021
		F08	Hospitalization or Prolonged Hospitalization	A0705	Battery Problem	05/11/2021	05/11/2021
				A070504	Premature Discharge of Battery	05/11/2021	05/11/2021

							F1905	Device Revision or Replacement	A0705	Battery Problem	05/11/2021	05/11/2021
									A070504	Premature Discharge of Battery	05/11/2021	05/11/2021
					CT321	Long-term surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem	05/11/2021	05/11/2021
									A070504	Premature Discharge of Battery	05/11/2021	05/11/2021
							F08	Hospitalization or Prolonged Hospitalization	A0705	Battery Problem	05/11/2021	05/11/2021
									A070504	Premature Discharge of Battery	05/11/2021	05/11/2021
							F1905	Device Revision or Replacement	A0705	Battery Problem	05/11/2021	05/11/2021
									A070504	Premature Discharge of Battery	05/11/2021	05/11/2021
					CT449	Generators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem	05/11/2021	05/11/2021
									A070504	Premature Discharge of Battery	05/11/2021	05/11/2021
							F08	Hospitalization or Prolonged Hospitalization	A0705	Battery Problem	05/11/2021	05/11/2021
									A070504	Premature Discharge of Battery	05/11/2021	05/11/2021
							F1905	Device Revision or Replacement	A0705	Battery Problem	05/11/2021	05/11/2021
									A070504	Premature Discharge of Battery	05/11/2021	05/11/2021
					CT981	Single use	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem	05/11/2021	05/11/2021
									A070504	Premature Discharge of Battery	05/11/2021	05/11/2021
							F08	Hospitalization or Prolonged Hospitalization	A0705	Battery Problem	05/11/2021	05/11/2021
									A070504	Premature Discharge of Battery	05/11/2021	05/11/2021
							F1905	Device Revision or Replacement	A0705	Battery Problem	05/11/2021	05/11/2021
									A070504	Premature Discharge of Battery	05/11/2021	05/11/2021
					CT983	Surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem	05/11/2021	05/11/2021
									A070504	Premature Discharge of Battery	05/11/2021	05/11/2021
							F08	Hospitalization or Prolonged Hospitalization	A0705	Battery Problem	05/11/2021	05/11/2021
									A070504	Premature Discharge of Battery	05/11/2021	05/11/2021
							F1905	Device Revision or Replacement	A0705	Battery Problem	05/11/2021	05/11/2021
									A070504	Premature Discharge of Battery	05/11/2021	05/11/2021
6854	2022	All other reportable incidents	47265	Dual-chamber implantable pacemaker, rate-responsive	CT141	Electrodes/Leads and associated devices	E0611	Heart Failure/Congestive Heart Failure	A070504	Premature Discharge of Battery	26/05/2022	26/05/2022
							E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	26/05/2022	26/05/2022
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	26/05/2022	26/05/2022
							F19	Surgical Intervention	A070504	Premature Discharge of Battery	26/05/2022	26/05/2022
					CT145	Cardiovascular devices	E0611	Heart Failure/Congestive Heart Failure	A070504	Premature Discharge of Battery	26/05/2022	26/05/2022
							E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	26/05/2022	26/05/2022
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	26/05/2022	26/05/2022
							F19	Surgical Intervention	A070504	Premature Discharge of Battery	26/05/2022	26/05/2022
					CT1684	Cardiac pacemakers and associated devices	E0611	Heart Failure/Congestive Heart Failure	A070504	Premature Discharge of Battery	26/05/2022	26/05/2022
							E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	26/05/2022	26/05/2022
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	26/05/2022	26/05/2022
							F19	Surgical Intervention	A070504	Premature Discharge of Battery	26/05/2022	26/05/2022
					CT1685	Cardiac pacemakers	E0611	Heart Failure/Congestive Heart Failure	A070504	Premature Discharge of Battery	26/05/2022	26/05/2022
							E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	26/05/2022	26/05/2022
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	26/05/2022	26/05/2022
							F19	Surgical Intervention	A070504	Premature Discharge of Battery	26/05/2022	26/05/2022
					CT233	Surgical	E0611	Heart Failure/Congestive Heart Failure	A070504	Premature Discharge of Battery	26/05/2022	26/05/2022
							E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	26/05/2022	26/05/2022
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	26/05/2022	26/05/2022
							F19	Surgical Intervention	A070504	Premature Discharge of Battery	26/05/2022	26/05/2022
					CT2406	Implantable	E0611	Heart Failure/Congestive Heart Failure	A070504	Premature Discharge of Battery	26/05/2022	26/05/2022
							E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	26/05/2022	26/05/2022
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	26/05/2022	26/05/2022
							F19	Surgical Intervention	A070504	Premature Discharge of Battery	26/05/2022	26/05/2022
					CT2407	Active implantable	E0611	Heart Failure/Congestive Heart Failure	A070504	Premature Discharge of Battery	26/05/2022	26/05/2022
							E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	26/05/2022	26/05/2022
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	26/05/2022	26/05/2022
							F19	Surgical Intervention	A070504	Premature Discharge of Battery	26/05/2022	26/05/2022
					CT257	Implantable cardiac pacemakers	E0611	Heart Failure/Congestive Heart Failure	A070504	Premature Discharge of Battery	26/05/2022	26/05/2022
							E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	26/05/2022	26/05/2022
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	26/05/2022	26/05/2022
							F19	Surgical Intervention	A070504	Premature Discharge of Battery	26/05/2022	26/05/2022
					CT2727	Active implant devices/ancillaries	E0611	Heart Failure/Congestive Heart Failure	A070504	Premature Discharge of Battery	26/05/2022	26/05/2022
							E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	26/05/2022	26/05/2022
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	26/05/2022	26/05/2022
							F19	Surgical Intervention	A070504	Premature Discharge of Battery	26/05/2022	26/05/2022
					CT2750	Implantable cardiac electrical stimulators and associated devices	E0611	Heart Failure/Congestive Heart Failure	A070504	Premature Discharge of Battery	26/05/2022	26/05/2022
							E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	26/05/2022	26/05/2022
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	26/05/2022	26/05/2022
							F19	Surgical Intervention	A070504	Premature Discharge of Battery	26/05/2022	26/05/2022
					CT2866	Implantable cardiac electrical stimulators	E0611	Heart Failure/Congestive Heart Failure	A070504	Premature Discharge of Battery	26/05/2022	26/05/2022
							E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	26/05/2022	26/05/2022
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	26/05/2022	26/05/2022
							F19	Surgical Intervention	A070504	Premature Discharge of Battery	26/05/2022	26/05/2022
					CT2985	Active	E0611	Heart Failure/Congestive Heart Failure	A070504	Premature Discharge of Battery	26/05/2022	26/05/2022
							E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	26/05/2022	26/05/2022
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	26/05/2022	26/05/2022
							F19	Surgical Intervention	A070504	Premature Discharge of Battery	26/05/2022	26/05/2022
					CT3001	Implantable metal	E0611	Heart Failure/Congestive Heart Failure	A070504	Premature Discharge of Battery	26/05/2022	26/05/2022
							E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	26/05/2022	26/05/2022
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	26/05/2022	26/05/2022
							F19	Surgical Intervention	A070504	Premature Discharge of Battery	26/05/2022	26/05/2022
					CT3002	Implantable synthetic polymer	E0611	Heart Failure/Congestive Heart Failure	A070504	Premature Discharge of Battery	26/05/2022	26/05/2022
							E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	26/05/2022	26/05/2022
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	26/05/2022	26/05/2022
							F19	Surgical Intervention	A070504	Premature Discharge of Battery	26/05/2022	26/05/2022
					CT3006	Implantable inorganic material	E0611	Heart Failure/Congestive Heart Failure	A070504	Premature Discharge of Battery	26/05/2022	26/05/2022
							E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	26/05/2022	26/05/2022
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	26/05/2022	26/05/2022
							F19	Surgical Intervention	A070504	Premature Discharge of Battery	26/05/2022	26/05/2022
					CT314	Electrically powered	E0611	Heart Failure/Congestive Heart Failure	A070504	Premature Discharge of Battery	26/05/2022	26/05/2022
							E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	26/05/2022	26/05/2022

							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	26/05/2022	26/05/2022
							F19	Surgical Intervention	A070504	Premature Discharge of Battery	26/05/2022	26/05/2022
					CT321	Long-term surgical invasive	E0611	Heart Failure/Congestive Heart Failure	A070504	Premature Discharge of Battery	26/05/2022	26/05/2022
							E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	26/05/2022	26/05/2022
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	26/05/2022	26/05/2022
							F19	Surgical Intervention	A070504	Premature Discharge of Battery	26/05/2022	26/05/2022
					CT449	Generators and associated devices	E0611	Heart Failure/Congestive Heart Failure	A070504	Premature Discharge of Battery	26/05/2022	26/05/2022
							E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	26/05/2022	26/05/2022
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	26/05/2022	26/05/2022
							F19	Surgical Intervention	A070504	Premature Discharge of Battery	26/05/2022	26/05/2022
					CT981	Single use	E0611	Heart Failure/Congestive Heart Failure	A070504	Premature Discharge of Battery	26/05/2022	26/05/2022
							E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	26/05/2022	26/05/2022
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	26/05/2022	26/05/2022
							F19	Surgical Intervention	A070504	Premature Discharge of Battery	26/05/2022	26/05/2022
					CT983	Surgical invasive	E0611	Heart Failure/Congestive Heart Failure	A070504	Premature Discharge of Battery	26/05/2022	26/05/2022
							E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	26/05/2022	26/05/2022
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	26/05/2022	26/05/2022
							F19	Surgical Intervention	A070504	Premature Discharge of Battery	26/05/2022	26/05/2022
6855	2022	All other reportable incidents	47265	Dual-chamber implantable pacemaker, rate-responsive	CT141	Electrodes/Leads and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT145	Cardiovascular devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT233	Surgical	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT2406	Implantable	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT2407	Active implantable	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT2727	Active implant devices/ancillaries	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT2750	Implantable cardiac electrical stimulators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT2985	Active	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT3001	Implantable metal	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT3002	Implantable synthetic polymer	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT3006	Implantable inorganic material	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT314	Electrically powered	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT321	Long-term surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT449	Generators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT981	Single use	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT983	Surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
6856	2022	All other reportable incidents	35223	Endocardial/interventricular septal pacing lead	CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070909	Oversensing		
									A090801	Signal Artifact/Noise		
							F26	No Health Consequences or Impact	A070909	Oversensing		
									A090801	Signal Artifact/Noise		
6857	2022	All other reportable incidents	35852	Single-chamber implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem	21/09/2022	21/09/2022
									A070504	Premature Discharge of Battery	21/09/2022	21/09/2022
							F08	Hospitalization or Prolonged Hospitalization	A0705	Battery Problem	21/09/2022	21/09/2022
									A070504	Premature Discharge of Battery	21/09/2022	21/09/2022
							F1905	Device Revision or Replacement	A0705	Battery Problem	21/09/2022	21/09/2022
									A070504	Premature Discharge of Battery	21/09/2022	21/09/2022
6858	2022	All other reportable incidents	17882	Physiologic-monitoring defibrillation system	CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A07	Electrical /Electronic Property Problem		
							F27	Problem Identified During Non-Clinical Procedure	A07	Electrical /Electronic Property Problem		
6859	2022	All other reportable incidents	47263	Cardiac resynchronization therapy implantable pacemaker	CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A0722	Impedance Problem		
									A1301	Failure to Read Input Signal		
							F26	No Health Consequences or Impact	A0722	Impedance Problem		
									A1301	Failure to Read Input Signal		
					CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A0722	Impedance Problem		
									A1301	Failure to Read Input Signal		
							F26	No Health Consequences or Impact	A0722	Impedance Problem		
									A1301	Failure to Read Input Signal		
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A0722	Impedance Problem		
									A1301	Failure to Read Input Signal		
							F26	No Health Consequences or Impact	A0722	Impedance Problem		
									A1301	Failure to Read Input Signal		
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A0722	Impedance Problem		
									A1301	Failure to Read Input Signal		
							F26	No Health Consequences or Impact	A0722	Impedance Problem		
									A1301	Failure to Read Input Signal		
6860	2022	All other reportable incidents	17882	Physiologic-monitoring defibrillation system	CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A071301	Failure to Deliver Shock/Stimulation		
							F27	Problem Identified During Non-Clinical Procedure	A071301	Failure to Deliver Shock/Stimulation		

6861	2022	All other reportable incidents	35852	Single-chamber implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E2104	Electric Shock	A070909	Oversensing		
									A071302	Inappropriate/Inadequate Shock/Stimulation		
									A090801	Signal Artifact/Noise		
									A1301	Failure to Read Input Signal		
							E2403	No Clinical Signs, Symptoms or Conditions	A070909	Oversensing		
									A071302	Inappropriate/Inadequate Shock/Stimulation		
									A090801	Signal Artifact/Noise		
									A1301	Failure to Read Input Signal		
							F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A070909	Oversensing		
									A071302	Inappropriate/Inadequate Shock/Stimulation		
									A090801	Signal Artifact/Noise		
									A1301	Failure to Read Input Signal		
6862	2022	All other reportable incidents	36241	Pacemaker/defibrillator lead	CT1684	Cardiac pacemakers and associated devices	E2311	Discomfort	A010402	Migration		
									A051201	Device Dislodged or Dislocated		
							E2403	No Clinical Signs, Symptoms or Conditions	A010402	Migration		
									A051201	Device Dislodged or Dislocated		
							F2203	Imaging Required	A010402	Migration		
									A051201	Device Dislodged or Dislocated		
							F26	No Health Consequences or Impact	A010402	Migration		
									A051201	Device Dislodged or Dislocated		
6863	2022	All other reportable incidents	60789	Intracardiac pacemaker	CT130	Catheters and associated devices	E0514	Thrombosis/Thrombus	A070101	Failure to Capture		
									A070102	High Capture Threshold		
									A072201	High Impedance		
									A1502	Positioning Problem		
							E0717	Dyspnea	A070101	Failure to Capture		
									A070102	High Capture Threshold		
									A072201	High Impedance		
									A1502	Positioning Problem		
							E233001	Chest Pain	A070101	Failure to Capture		
									A070102	High Capture Threshold		
									A072201	High Impedance		
									A1502	Positioning Problem		
							E2403	No Clinical Signs, Symptoms or Conditions	A070101	Failure to Capture		
									A070102	High Capture Threshold		
									A072201	High Impedance		
									A1502	Positioning Problem		
							F08	Hospitalization or Prolonged Hospitalization	A070101	Failure to Capture		
									A070102	High Capture Threshold		
									A072201	High Impedance		
									A1502	Positioning Problem		
							F26	No Health Consequences or Impact	A070101	Failure to Capture		
									A070102	High Capture Threshold		
									A072201	High Impedance		
									A1502	Positioning Problem		
					CT145	Cardiovascular devices	E0514	Thrombosis/Thrombus	A070101	Failure to Capture		
									A070102	High Capture Threshold		
									A072201	High Impedance		
									A1502	Positioning Problem		
							E0717	Dyspnea	A070101	Failure to Capture		
									A070102	High Capture Threshold		
									A072201	High Impedance		
									A1502	Positioning Problem		
							E233001	Chest Pain	A070101	Failure to Capture		
									A070102	High Capture Threshold		
									A072201	High Impedance		
									A1502	Positioning Problem		
							E2403	No Clinical Signs, Symptoms or Conditions	A070101	Failure to Capture		
									A070102	High Capture Threshold		
									A072201	High Impedance		
									A1502	Positioning Problem		
							F08	Hospitalization or Prolonged Hospitalization	A070101	Failure to Capture		
									A070102	High Capture Threshold		
									A072201	High Impedance		
									A1502	Positioning Problem		
							F26	No Health Consequences or Impact	A070101	Failure to Capture		
									A070102	High Capture Threshold		
									A072201	High Impedance		
									A1502	Positioning Problem		
					CT1684	Cardiac pacemakers and associated devices	E0514	Thrombosis/Thrombus	A070101	Failure to Capture		
									A070102	High Capture Threshold		
									A072201	High Impedance		
									A1502	Positioning Problem		
							E0717	Dyspnea	A070101	Failure to Capture		
									A070102	High Capture Threshold		
									A072201	High Impedance		
									A1502	Positioning Problem		
							E233001	Chest Pain	A070101	Failure to Capture		
									A070102	High Capture Threshold		
									A072201	High Impedance		
									A1502	Positioning Problem		
							E2403	No Clinical Signs, Symptoms or Conditions	A070101	Failure to Capture		
									A070102	High Capture Threshold		
									A072201	High Impedance		
									A1502	Positioning Problem		
							F08	Hospitalization or Prolonged Hospitalization	A070101	Failure to Capture		
									A070102	High Capture Threshold		
									A072201	High Impedance		
									A1502	Positioning Problem		

							F26	No Health Consequences or Impact	A070101	Failure to Capture		
									A070102	High Capture Threshold		
									A072201	High Impedance		
									A1502	Positioning Problem		
					CT1685	Cardiac pacemakers	E0514	Thrombosis/Thrombus	A070101	Failure to Capture		
									A070102	High Capture Threshold		
									A072201	High Impedance		
									A1502	Positioning Problem		
							E0717	Dyspnea	A070101	Failure to Capture		
									A070102	High Capture Threshold		
									A072201	High Impedance		
									A1502	Positioning Problem		
							E233001	Chest Pain	A070101	Failure to Capture		
									A070102	High Capture Threshold		
									A072201	High Impedance		
									A1502	Positioning Problem		
							E2403	No Clinical Signs, Symptoms or Conditions	A070101	Failure to Capture		
									A070102	High Capture Threshold		
									A072201	High Impedance		
									A1502	Positioning Problem		
							F08	Hospitalization or Prolonged Hospitalization	A070101	Failure to Capture		
									A070102	High Capture Threshold		
									A072201	High Impedance		
									A1502	Positioning Problem		
							F26	No Health Consequences or Impact	A070101	Failure to Capture		
									A070102	High Capture Threshold		
									A072201	High Impedance		
									A1502	Positioning Problem		
					CT233	Surgical	E0514	Thrombosis/Thrombus	A070101	Failure to Capture		
									A070102	High Capture Threshold		
									A072201	High Impedance		
									A1502	Positioning Problem		
							E0717	Dyspnea	A070101	Failure to Capture		
									A070102	High Capture Threshold		
									A072201	High Impedance		
									A1502	Positioning Problem		
							E233001	Chest Pain	A070101	Failure to Capture		
									A070102	High Capture Threshold		
									A072201	High Impedance		
									A1502	Positioning Problem		
							E2403	No Clinical Signs, Symptoms or Conditions	A070101	Failure to Capture		
									A070102	High Capture Threshold		
									A072201	High Impedance		
									A1502	Positioning Problem		
							F08	Hospitalization or Prolonged Hospitalization	A070101	Failure to Capture		
									A070102	High Capture Threshold		
									A072201	High Impedance		
									A1502	Positioning Problem		
							F26	No Health Consequences or Impact	A070101	Failure to Capture		
									A070102	High Capture Threshold		
									A072201	High Impedance		
									A1502	Positioning Problem		
					CT2406	Implantable	E0514	Thrombosis/Thrombus	A070101	Failure to Capture		
									A070102	High Capture Threshold		
									A072201	High Impedance		
									A1502	Positioning Problem		
							E0717	Dyspnea	A070101	Failure to Capture		
									A070102	High Capture Threshold		
									A072201	High Impedance		
									A1502	Positioning Problem		
							E233001	Chest Pain	A070101	Failure to Capture		
									A070102	High Capture Threshold		
									A072201	High Impedance		
									A1502	Positioning Problem		
							E2403	No Clinical Signs, Symptoms or Conditions	A070101	Failure to Capture		
									A070102	High Capture Threshold		
									A072201	High Impedance		
									A1502	Positioning Problem		
							F08	Hospitalization or Prolonged Hospitalization	A070101	Failure to Capture		
									A070102	High Capture Threshold		
									A072201	High Impedance		
									A1502	Positioning Problem		
							F26	No Health Consequences or Impact	A070101	Failure to Capture		
									A070102	High Capture Threshold		
									A072201	High Impedance		
									A1502	Positioning Problem		
					CT2407	Active implantable	E0514	Thrombosis/Thrombus	A070101	Failure to Capture		
									A070102	High Capture Threshold		
									A072201	High Impedance		
									A1502	Positioning Problem		
							E0717	Dyspnea	A070101	Failure to Capture		
									A070102	High Capture Threshold		
									A072201	High Impedance		
									A1502	Positioning Problem		
							E233001	Chest Pain	A070101	Failure to Capture		
									A070102	High Capture Threshold		
									A072201	High Impedance		
									A1502	Positioning Problem		

						E2403	No Clinical Signs, Symptoms or Conditions	A070101	Failure to Capture		
								A070102	High Capture Threshold		
								A072201	High Impedance		
								A1502	Positioning Problem		
						F08	Hospitalization or Prolonged Hospitalization	A070101	Failure to Capture		
								A070102	High Capture Threshold		
								A072201	High Impedance		
								A1502	Positioning Problem		
						F26	No Health Consequences or Impact	A070101	Failure to Capture		
								A070102	High Capture Threshold		
								A072201	High Impedance		
								A1502	Positioning Problem		
					CT257	Implantable cardiac pacemakers	E0514	Thrombosis/Thrombus	A070101	Failure to Capture		
									A070102	High Capture Threshold		
									A072201	High Impedance		
									A1502	Positioning Problem		
							E0717	Dyspnea	A070101	Failure to Capture		
									A070102	High Capture Threshold		
									A072201	High Impedance		
									A1502	Positioning Problem		
							E233001	Chest Pain	A070101	Failure to Capture		
									A070102	High Capture Threshold		
									A072201	High Impedance		
									A1502	Positioning Problem		
							E2403	No Clinical Signs, Symptoms or Conditions	A070101	Failure to Capture		
									A070102	High Capture Threshold		
									A072201	High Impedance		
									A1502	Positioning Problem		
							F08	Hospitalization or Prolonged Hospitalization	A070101	Failure to Capture		
									A070102	High Capture Threshold		
									A072201	High Impedance		
									A1502	Positioning Problem		
							F26	No Health Consequences or Impact	A070101	Failure to Capture		
									A070102	High Capture Threshold		
									A072201	High Impedance		
									A1502	Positioning Problem		
					CT2727	Active implant devices/ancillaries	E0514	Thrombosis/Thrombus	A070101	Failure to Capture		
									A070102	High Capture Threshold		
									A072201	High Impedance		
									A1502	Positioning Problem		
							E0717	Dyspnea	A070101	Failure to Capture		
									A070102	High Capture Threshold		
									A072201	High Impedance		
									A1502	Positioning Problem		
							E233001	Chest Pain	A070101	Failure to Capture		
									A070102	High Capture Threshold		
									A072201	High Impedance		
									A1502	Positioning Problem		
							E2403	No Clinical Signs, Symptoms or Conditions	A070101	Failure to Capture		
									A070102	High Capture Threshold		
									A072201	High Impedance		
									A1502	Positioning Problem		
							F08	Hospitalization or Prolonged Hospitalization	A070101	Failure to Capture		
									A070102	High Capture Threshold		
									A072201	High Impedance		
									A1502	Positioning Problem		
							F26	No Health Consequences or Impact	A070101	Failure to Capture		
									A070102	High Capture Threshold		
									A072201	High Impedance		
									A1502	Positioning Problem		
					CT2750	Implantable cardiac electrical stimulators and associated devices	E0514	Thrombosis/Thrombus	A070101	Failure to Capture		
									A070102	High Capture Threshold		
									A072201	High Impedance		
									A1502	Positioning Problem		
							E0717	Dyspnea	A070101	Failure to Capture		
									A070102	High Capture Threshold		
									A072201	High Impedance		
									A1502	Positioning Problem		
							E233001	Chest Pain	A070101	Failure to Capture		
									A070102	High Capture Threshold		
									A072201	High Impedance		
									A1502	Positioning Problem		
							E2403	No Clinical Signs, Symptoms or Conditions	A070101	Failure to Capture		
									A070102	High Capture Threshold		
									A072201	High Impedance		
									A1502	Positioning Problem		
							F08	Hospitalization or Prolonged Hospitalization	A070101	Failure to Capture		
									A070102	High Capture Threshold		
									A072201	High Impedance		
									A1502	Positioning Problem		
							F26	No Health Consequences or Impact	A070101	Failure to Capture		
									A070102	High Capture Threshold		
									A072201	High Impedance		
									A1502	Positioning Problem		
					CT2818	Catheterization/Cannulation/Fluid path devices	E0514	Thrombosis/Thrombus	A070101	Failure to Capture		
									A070102	High Capture Threshold		
									A072201	High Impedance		
									A1502	Positioning Problem		

							E0717	Dyspnea	A070101	Failure to Capture		
									A070102	High Capture Threshold		
									A072201	High Impedance		
									A1502	Positioning Problem		
							E233001	Chest Pain	A070101	Failure to Capture		
									A070102	High Capture Threshold		
									A072201	High Impedance		
									A1502	Positioning Problem		
							E2403	No Clinical Signs, Symptoms or Conditions	A070101	Failure to Capture		
									A070102	High Capture Threshold		
									A072201	High Impedance		
									A1502	Positioning Problem		
							F08	Hospitalization or Prolonged Hospitalization	A070101	Failure to Capture		
									A070102	High Capture Threshold		
									A072201	High Impedance		
									A1502	Positioning Problem		
							F26	No Health Consequences or Impact	A070101	Failure to Capture		
									A070102	High Capture Threshold		
									A072201	High Impedance		
									A1502	Positioning Problem		
						CT2866	Implantable cardiac electrical stimulators	E0514	Thrombosis/Thrombus	A070101	Failure to Capture		
										A070102	High Capture Threshold		
										A072201	High Impedance		
										A1502	Positioning Problem		
								E0717	Dyspnea	A070101	Failure to Capture		
										A070102	High Capture Threshold		
										A072201	High Impedance		
										A1502	Positioning Problem		
								E233001	Chest Pain	A070101	Failure to Capture		
										A070102	High Capture Threshold		
										A072201	High Impedance		
										A1502	Positioning Problem		
								E2403	No Clinical Signs, Symptoms or Conditions	A070101	Failure to Capture		
										A070102	High Capture Threshold		
										A072201	High Impedance		
										A1502	Positioning Problem		
								F08	Hospitalization or Prolonged Hospitalization	A070101	Failure to Capture		
										A070102	High Capture Threshold		
										A072201	High Impedance		
										A1502	Positioning Problem		
								F26	No Health Consequences or Impact	A070101	Failure to Capture		
										A070102	High Capture Threshold		
										A072201	High Impedance		
										A1502	Positioning Problem		
						CT2985	Active	E0514	Thrombosis/Thrombus	A070101	Failure to Capture		
										A070102	High Capture Threshold		
										A072201	High Impedance		
										A1502	Positioning Problem		
								E0717	Dyspnea	A070101	Failure to Capture		
										A070102	High Capture Threshold		
										A072201	High Impedance		
										A1502	Positioning Problem		
								E233001	Chest Pain	A070101	Failure to Capture		
										A070102	High Capture Threshold		
										A072201	High Impedance		
										A1502	Positioning Problem		
								E2403	No Clinical Signs, Symptoms or Conditions	A070101	Failure to Capture		
										A070102	High Capture Threshold		
										A072201	High Impedance		
										A1502	Positioning Problem		
								F08	Hospitalization or Prolonged Hospitalization	A070101	Failure to Capture		
										A070102	High Capture Threshold		
										A072201	High Impedance		
										A1502	Positioning Problem		
								F26	No Health Consequences or Impact	A070101	Failure to Capture		
										A070102	High Capture Threshold		
										A072201	High Impedance		
										A1502	Positioning Problem		
						CT3001	Implantable metal	E0514	Thrombosis/Thrombus	A070101	Failure to Capture		
										A070102	High Capture Threshold		
										A072201	High Impedance		
										A1502	Positioning Problem		
								E0717	Dyspnea	A070101	Failure to Capture		
										A070102	High Capture Threshold		
										A072201	High Impedance		
										A1502	Positioning Problem		
								E233001	Chest Pain	A070101	Failure to Capture		
										A070102	High Capture Threshold		
										A072201	High Impedance		
										A1502	Positioning Problem		
								E2403	No Clinical Signs, Symptoms or Conditions	A070101	Failure to Capture		
										A070102	High Capture Threshold		
										A072201	High Impedance		
										A1502	Positioning Problem		
								F08	Hospitalization or Prolonged Hospitalization	A070101	Failure to Capture		
										A070102	High Capture Threshold		
										A072201	High Impedance		
										A1502	Positioning Problem		

							F26	No Health Consequences or Impact	A070101	Failure to Capture		
									A070102	High Capture Threshold		
									A072201	High Impedance		
									A1502	Positioning Problem		
					CT3002	Implantable synthetic polymer	E0514	Thrombosis/Thrombus	A070101	Failure to Capture		
									A070102	High Capture Threshold		
									A072201	High Impedance		
									A1502	Positioning Problem		
							E0717	Dyspnea	A070101	Failure to Capture		
									A070102	High Capture Threshold		
									A072201	High Impedance		
									A1502	Positioning Problem		
							E233001	Chest Pain	A070101	Failure to Capture		
									A070102	High Capture Threshold		
									A072201	High Impedance		
									A1502	Positioning Problem		
							E2403	No Clinical Signs, Symptoms or Conditions	A070101	Failure to Capture		
									A070102	High Capture Threshold		
									A072201	High Impedance		
									A1502	Positioning Problem		
							F08	Hospitalization or Prolonged Hospitalization	A070101	Failure to Capture		
									A070102	High Capture Threshold		
									A072201	High Impedance		
									A1502	Positioning Problem		
							F26	No Health Consequences or Impact	A070101	Failure to Capture		
									A070102	High Capture Threshold		
									A072201	High Impedance		
									A1502	Positioning Problem		
					CT3006	Implantable inorganic material	E0514	Thrombosis/Thrombus	A070101	Failure to Capture		
									A070102	High Capture Threshold		
									A072201	High Impedance		
									A1502	Positioning Problem		
							E0717	Dyspnea	A070101	Failure to Capture		
									A070102	High Capture Threshold		
									A072201	High Impedance		
									A1502	Positioning Problem		
							E233001	Chest Pain	A070101	Failure to Capture		
									A070102	High Capture Threshold		
									A072201	High Impedance		
									A1502	Positioning Problem		
							E2403	No Clinical Signs, Symptoms or Conditions	A070101	Failure to Capture		
									A070102	High Capture Threshold		
									A072201	High Impedance		
									A1502	Positioning Problem		
							F08	Hospitalization or Prolonged Hospitalization	A070101	Failure to Capture		
									A070102	High Capture Threshold		
									A072201	High Impedance		
									A1502	Positioning Problem		
							F26	No Health Consequences or Impact	A070101	Failure to Capture		
									A070102	High Capture Threshold		
									A072201	High Impedance		
									A1502	Positioning Problem		
					CT314	Electrically powered	E0514	Thrombosis/Thrombus	A070101	Failure to Capture		
									A070102	High Capture Threshold		
									A072201	High Impedance		
									A1502	Positioning Problem		
							E0717	Dyspnea	A070101	Failure to Capture		
									A070102	High Capture Threshold		
									A072201	High Impedance		
									A1502	Positioning Problem		
							E233001	Chest Pain	A070101	Failure to Capture		
									A070102	High Capture Threshold		
									A072201	High Impedance		
									A1502	Positioning Problem		
							E2403	No Clinical Signs, Symptoms or Conditions	A070101	Failure to Capture		
									A070102	High Capture Threshold		
									A072201	High Impedance		
									A1502	Positioning Problem		
							F08	Hospitalization or Prolonged Hospitalization	A070101	Failure to Capture		
									A070102	High Capture Threshold		
									A072201	High Impedance		
									A1502	Positioning Problem		
							F26	No Health Consequences or Impact	A070101	Failure to Capture		
									A070102	High Capture Threshold		
									A072201	High Impedance		
									A1502	Positioning Problem		
					CT321	Long-term surgical invasive	E0514	Thrombosis/Thrombus	A070101	Failure to Capture		
									A070102	High Capture Threshold		
									A072201	High Impedance		
									A1502	Positioning Problem		
							E0717	Dyspnea	A070101	Failure to Capture		
									A070102	High Capture Threshold		
									A072201	High Impedance		
									A1502	Positioning Problem		
							E233001	Chest Pain	A070101	Failure to Capture		
									A070102	High Capture Threshold		
									A072201	High Impedance		
									A1502	Positioning Problem		

							E2403	No Clinical Signs, Symptoms or Conditions	A070101	Failure to Capture		
									A070102	High Capture Threshold		
									A072201	High Impedance		
									A1502	Positioning Problem		
							F08	Hospitalization or Prolonged Hospitalization	A070101	Failure to Capture		
									A070102	High Capture Threshold		
									A072201	High Impedance		
									A1502	Positioning Problem		
							F26	No Health Consequences or Impact	A070101	Failure to Capture		
									A070102	High Capture Threshold		
									A072201	High Impedance		
									A1502	Positioning Problem		
					CT981	Single use	E0514	Thrombosis/Thrombus	A070101	Failure to Capture		
									A070102	High Capture Threshold		
									A072201	High Impedance		
									A1502	Positioning Problem		
							E0717	Dyspnea	A070101	Failure to Capture		
									A070102	High Capture Threshold		
									A072201	High Impedance		
									A1502	Positioning Problem		
							E233001	Chest Pain	A070101	Failure to Capture		
									A070102	High Capture Threshold		
									A072201	High Impedance		
									A1502	Positioning Problem		
							E2403	No Clinical Signs, Symptoms or Conditions	A070101	Failure to Capture		
									A070102	High Capture Threshold		
									A072201	High Impedance		
									A1502	Positioning Problem		
							F08	Hospitalization or Prolonged Hospitalization	A070101	Failure to Capture		
									A070102	High Capture Threshold		
									A072201	High Impedance		
									A1502	Positioning Problem		
							F26	No Health Consequences or Impact	A070101	Failure to Capture		
									A070102	High Capture Threshold		
									A072201	High Impedance		
									A1502	Positioning Problem		
					CT983	Surgical invasive	E0514	Thrombosis/Thrombus	A070101	Failure to Capture		
									A070102	High Capture Threshold		
									A072201	High Impedance		
									A1502	Positioning Problem		
							E0717	Dyspnea	A070101	Failure to Capture		
									A070102	High Capture Threshold		
									A072201	High Impedance		
									A1502	Positioning Problem		
							E233001	Chest Pain	A070101	Failure to Capture		
									A070102	High Capture Threshold		
									A072201	High Impedance		
									A1502	Positioning Problem		
							E2403	No Clinical Signs, Symptoms or Conditions	A070101	Failure to Capture		
									A070102	High Capture Threshold		
									A072201	High Impedance		
									A1502	Positioning Problem		
							F08	Hospitalization or Prolonged Hospitalization	A070101	Failure to Capture		
									A070102	High Capture Threshold		
									A072201	High Impedance		
									A1502	Positioning Problem		
							F26	No Health Consequences or Impact	A070101	Failure to Capture		
									A070102	High Capture Threshold		
									A072201	High Impedance		
									A1502	Positioning Problem		
6864	2021	All other reportable incidents	37265	Dual-chamber implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
6865	2021	All other reportable incidents	47270	Cardiac resynchronization therapy implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A072201	High Impedance		
							F26	No Health Consequences or Impact	A072201	High Impedance		
6866	2021	All other reportable incidents	36241	Pacemaker/defibrillator lead	CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A072202	Low Impedance		
							F26	No Health Consequences or Impact	A072202	Low Impedance		
6867	2022	All other reportable incidents	47263	Cardiac resynchronization therapy implantable pacemaker	CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A1107	Data Problem	12/07/2022	12/07/2022
							F08	Hospitalization or Prolonged Hospitalization	A1107	Data Problem	12/07/2022	12/07/2022
							F1903	Device Explantation	A1107	Data Problem	12/07/2022	12/07/2022
					CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A1107	Data Problem	12/07/2022	12/07/2022
							F08	Hospitalization or Prolonged Hospitalization	A1107	Data Problem	12/07/2022	12/07/2022
							F1903	Device Explantation	A1107	Data Problem	12/07/2022	12/07/2022
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A1107	Data Problem	12/07/2022	12/07/2022
							F08	Hospitalization or Prolonged Hospitalization	A1107	Data Problem	12/07/2022	12/07/2022
							F1903	Device Explantation	A1107	Data Problem	12/07/2022	12/07/2022
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A1107	Data Problem	12/07/2022	12/07/2022
							F08	Hospitalization or Prolonged Hospitalization	A1107	Data Problem	12/07/2022	12/07/2022
							F1903	Device Explantation	A1107	Data Problem	12/07/2022	12/07/2022
6868	2022	All other reportable incidents	17882	Physiologic-monitoring defibrillation system	CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070908	Failure to Sense		
							F27	Problem Identified During Non-Clinical Procedure	A070908	Failure to Sense		
6869	2022	All other reportable incidents	47263	Cardiac resynchronization therapy implantable pacemaker	CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		

					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
6870	2021	All other reportable incidents	47267	Single-chamber implantable pacemaker, rate-responsive	CT141	Electrodes/Leads and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	06/02/2022	06/02/2022
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	06/02/2022	06/02/2022
							F1903	Device Explantation	A070504	Premature Discharge of Battery	06/02/2022	06/02/2022
					CT145	Cardiovascular devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	06/02/2022	06/02/2022
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	06/02/2022	06/02/2022
							F1903	Device Explantation	A070504	Premature Discharge of Battery	06/02/2022	06/02/2022
					CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	06/02/2022	06/02/2022
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	06/02/2022	06/02/2022
							F1903	Device Explantation	A070504	Premature Discharge of Battery	06/02/2022	06/02/2022
					CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	06/02/2022	06/02/2022
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	06/02/2022	06/02/2022
							F1903	Device Explantation	A070504	Premature Discharge of Battery	06/02/2022	06/02/2022
					CT233	Surgical	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	06/02/2022	06/02/2022
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	06/02/2022	06/02/2022
							F1903	Device Explantation	A070504	Premature Discharge of Battery	06/02/2022	06/02/2022
					CT2406	Implantable	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	06/02/2022	06/02/2022
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	06/02/2022	06/02/2022
							F1903	Device Explantation	A070504	Premature Discharge of Battery	06/02/2022	06/02/2022
					CT2407	Active implantable	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	06/02/2022	06/02/2022
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	06/02/2022	06/02/2022
							F1903	Device Explantation	A070504	Premature Discharge of Battery	06/02/2022	06/02/2022
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	06/02/2022	06/02/2022
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	06/02/2022	06/02/2022
							F1903	Device Explantation	A070504	Premature Discharge of Battery	06/02/2022	06/02/2022
					CT2727	Active implant devices/ancillaries	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	06/02/2022	06/02/2022
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	06/02/2022	06/02/2022
							F1903	Device Explantation	A070504	Premature Discharge of Battery	06/02/2022	06/02/2022
					CT2750	Implantable cardiac electrical stimulators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	06/02/2022	06/02/2022
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	06/02/2022	06/02/2022
							F1903	Device Explantation	A070504	Premature Discharge of Battery	06/02/2022	06/02/2022
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	06/02/2022	06/02/2022
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	06/02/2022	06/02/2022
							F1903	Device Explantation	A070504	Premature Discharge of Battery	06/02/2022	06/02/2022
					CT2985	Active	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	06/02/2022	06/02/2022
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	06/02/2022	06/02/2022
							F1903	Device Explantation	A070504	Premature Discharge of Battery	06/02/2022	06/02/2022
					CT3001	Implantable metal	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	06/02/2022	06/02/2022
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	06/02/2022	06/02/2022
							F1903	Device Explantation	A070504	Premature Discharge of Battery	06/02/2022	06/02/2022
					CT3002	Implantable synthetic polymer	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	06/02/2022	06/02/2022
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	06/02/2022	06/02/2022
							F1903	Device Explantation	A070504	Premature Discharge of Battery	06/02/2022	06/02/2022
					CT3006	Implantable inorganic material	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	06/02/2022	06/02/2022
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	06/02/2022	06/02/2022
							F1903	Device Explantation	A070504	Premature Discharge of Battery	06/02/2022	06/02/2022
					CT314	Electrically powered	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	06/02/2022	06/02/2022
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	06/02/2022	06/02/2022
							F1903	Device Explantation	A070504	Premature Discharge of Battery	06/02/2022	06/02/2022
					CT321	Long-term surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	06/02/2022	06/02/2022
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	06/02/2022	06/02/2022
							F1903	Device Explantation	A070504	Premature Discharge of Battery	06/02/2022	06/02/2022
					CT449	Generators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	06/02/2022	06/02/2022
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	06/02/2022	06/02/2022
							F1903	Device Explantation	A070504	Premature Discharge of Battery	06/02/2022	06/02/2022
					CT981	Single use	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	06/02/2022	06/02/2022
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	06/02/2022	06/02/2022
							F1903	Device Explantation	A070504	Premature Discharge of Battery	06/02/2022	06/02/2022
					CT983	Surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	06/02/2022	06/02/2022
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	06/02/2022	06/02/2022
							F1903	Device Explantation	A070504	Premature Discharge of Battery	06/02/2022	06/02/2022
6871	2021	All other reportable incidents	36241	Pacemaker/defibrillator lead	CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A072201	High Impedance		
							F26	No Health Consequences or Impact	A072201	High Impedance		
6872	2021	All other reportable incidents	47270	Cardiac resynchronization therapy implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A070102	High Capture Threshold		
							F26	No Health Consequences or Impact	A070102	High Capture Threshold		
6873	2021	All other reportable incidents	17882	Physiologic-monitoring defibrillation system	CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A1603	Failure of Device to Self-Test		
							F27	Problem Identified During Non-Clinical Procedure	A1603	Failure of Device to Self-Test		
6874	2021	All other reportable incidents	36241	Pacemaker/defibrillator lead	CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A072202	Low Impedance		
							F26	No Health Consequences or Impact	A072202	Low Impedance		
6875	2021	All other reportable incidents	37265	Dual-chamber implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E2401	Insufficient Information	A160601	Premature Elective Replacement Indicator	23/03/2021	23/03/2021
							E2403	No Clinical Signs, Symptoms or Conditions	A160601	Premature Elective Replacement Indicator	23/03/2021	23/03/2021
							F1905	Device Revision or Replacement	A160601	Premature Elective Replacement Indicator	23/03/2021	23/03/2021
6876	2021	All other reportable incidents	47265	Dual-chamber implantable pacemaker, rate-responsive	CT141	Electrodes/Leads and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	15/10/2021	15/10/2021
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	15/10/2021	15/10/2021
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	15/10/2021	15/10/2021
					CT145	Cardiovascular devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	15/10/2021	15/10/2021
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	15/10/2021	15/10/2021
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	15/10/2021	15/10/2021
					CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	15/10/2021	15/10/2021
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	15/10/2021	15/10/2021
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	15/10/2021	15/10/2021
					CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	15/10/2021	15/10/2021
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	15/10/2021	15/10/2021
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	15/10/2021	15/10/2021
					CT233	Surgical	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	15/10/2021	15/10/2021
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	15/10/2021	15/10/2021
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	15/10/2021	15/10/2021

					CT2406	Implantable	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	15/10/2021	15/10/2021
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	15/10/2021	15/10/2021
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	15/10/2021	15/10/2021
					CT2407	Active implantable	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	15/10/2021	15/10/2021
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	15/10/2021	15/10/2021
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	15/10/2021	15/10/2021
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	15/10/2021	15/10/2021
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	15/10/2021	15/10/2021
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	15/10/2021	15/10/2021
					CT2727	Active implant devices/ancillaries	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	15/10/2021	15/10/2021
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	15/10/2021	15/10/2021
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	15/10/2021	15/10/2021
					CT2750	Implantable cardiac electrical stimulators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	15/10/2021	15/10/2021
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	15/10/2021	15/10/2021
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	15/10/2021	15/10/2021
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	15/10/2021	15/10/2021
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	15/10/2021	15/10/2021
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	15/10/2021	15/10/2021
					CT2985	Active	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	15/10/2021	15/10/2021
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	15/10/2021	15/10/2021
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	15/10/2021	15/10/2021
					CT3001	Implantable metal	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	15/10/2021	15/10/2021
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	15/10/2021	15/10/2021
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	15/10/2021	15/10/2021
					CT3002	Implantable synthetic polymer	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	15/10/2021	15/10/2021
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	15/10/2021	15/10/2021
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	15/10/2021	15/10/2021
					CT3006	Implantable inorganic material	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	15/10/2021	15/10/2021
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	15/10/2021	15/10/2021
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	15/10/2021	15/10/2021
					CT314	Electrically powered	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	15/10/2021	15/10/2021
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	15/10/2021	15/10/2021
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	15/10/2021	15/10/2021
					CT321	Long-term surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	15/10/2021	15/10/2021
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	15/10/2021	15/10/2021
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	15/10/2021	15/10/2021
					CT449	Generators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	15/10/2021	15/10/2021
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	15/10/2021	15/10/2021
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	15/10/2021	15/10/2021
					CT981	Single use	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	15/10/2021	15/10/2021
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	15/10/2021	15/10/2021
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	15/10/2021	15/10/2021
					CT983	Surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	15/10/2021	15/10/2021
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	15/10/2021	15/10/2021
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	15/10/2021	15/10/2021
6877	2021	All other reportable incidents	47267	Single-chamber implantable pacemaker, rate-responsive	CT141	Electrodes/Leads and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem	10/11/2021	10/11/2021
									A070504	Premature Discharge of Battery	10/11/2021	10/11/2021
							F08	Hospitalization or Prolonged Hospitalization	A0705	Battery Problem	10/11/2021	10/11/2021
									A070504	Premature Discharge of Battery	10/11/2021	10/11/2021
							F1905	Device Revision or Replacement	A0705	Battery Problem	10/11/2021	10/11/2021
									A070504	Premature Discharge of Battery	10/11/2021	10/11/2021
					CT145	Cardiovascular devices	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem	10/11/2021	10/11/2021
									A070504	Premature Discharge of Battery	10/11/2021	10/11/2021
							F08	Hospitalization or Prolonged Hospitalization	A0705	Battery Problem	10/11/2021	10/11/2021
									A070504	Premature Discharge of Battery	10/11/2021	10/11/2021
							F1905	Device Revision or Replacement	A0705	Battery Problem	10/11/2021	10/11/2021
									A070504	Premature Discharge of Battery	10/11/2021	10/11/2021
					CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem	10/11/2021	10/11/2021
									A070504	Premature Discharge of Battery	10/11/2021	10/11/2021
							F08	Hospitalization or Prolonged Hospitalization	A0705	Battery Problem	10/11/2021	10/11/2021
									A070504	Premature Discharge of Battery	10/11/2021	10/11/2021
							F1905	Device Revision or Replacement	A0705	Battery Problem	10/11/2021	10/11/2021
									A070504	Premature Discharge of Battery	10/11/2021	10/11/2021
					CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem	10/11/2021	10/11/2021
									A070504	Premature Discharge of Battery	10/11/2021	10/11/2021
							F08	Hospitalization or Prolonged Hospitalization	A0705	Battery Problem	10/11/2021	10/11/2021
									A070504	Premature Discharge of Battery	10/11/2021	10/11/2021
							F1905	Device Revision or Replacement	A0705	Battery Problem	10/11/2021	10/11/2021
									A070504	Premature Discharge of Battery	10/11/2021	10/11/2021
					CT233	Surgical	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem	10/11/2021	10/11/2021
									A070504	Premature Discharge of Battery	10/11/2021	10/11/2021
							F08	Hospitalization or Prolonged Hospitalization	A0705	Battery Problem	10/11/2021	10/11/2021
									A070504	Premature Discharge of Battery	10/11/2021	10/11/2021
							F1905	Device Revision or Replacement	A0705	Battery Problem	10/11/2021	10/11/2021
									A070504	Premature Discharge of Battery	10/11/2021	10/11/2021
					CT2406	Implantable	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem	10/11/2021	10/11/2021
									A070504	Premature Discharge of Battery	10/11/2021	10/11/2021
							F08	Hospitalization or Prolonged Hospitalization	A0705	Battery Problem	10/11/2021	10/11/2021
									A070504	Premature Discharge of Battery	10/11/2021	10/11/2021
							F1905	Device Revision or Replacement	A0705	Battery Problem	10/11/2021	10/11/2021
									A070504	Premature Discharge of Battery	10/11/2021	10/11/2021
					CT2407	Active implantable	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem	10/11/2021	10/11/2021
									A070504	Premature Discharge of Battery	10/11/2021	10/11/2021
							F08	Hospitalization or Prolonged Hospitalization	A0705	Battery Problem	10/11/2021	10/11/2021
									A070504	Premature Discharge of Battery	10/11/2021	10/11/2021
							F1905	Device Revision or Replacement	A0705	Battery Problem	10/11/2021	10/11/2021
									A070504	Premature Discharge of Battery	10/11/2021	10/11/2021
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem	10/11/2021	10/11/2021

									A070504	Premature Discharge of Battery	10/11/2021	10/11/2021
							F08	Hospitalization or Prolonged Hospitalization	A0705	Battery Problem	10/11/2021	10/11/2021
									A070504	Premature Discharge of Battery	10/11/2021	10/11/2021
							F1905	Device Revision or Replacement	A0705	Battery Problem	10/11/2021	10/11/2021
									A070504	Premature Discharge of Battery	10/11/2021	10/11/2021
					CT2727	Active implant devices/ancillaries	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem	10/11/2021	10/11/2021
									A070504	Premature Discharge of Battery	10/11/2021	10/11/2021
							F08	Hospitalization or Prolonged Hospitalization	A0705	Battery Problem	10/11/2021	10/11/2021
									A070504	Premature Discharge of Battery	10/11/2021	10/11/2021
							F1905	Device Revision or Replacement	A0705	Battery Problem	10/11/2021	10/11/2021
									A070504	Premature Discharge of Battery	10/11/2021	10/11/2021
					CT2750	Implantable cardiac electrical stimulators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem	10/11/2021	10/11/2021
									A070504	Premature Discharge of Battery	10/11/2021	10/11/2021
							F08	Hospitalization or Prolonged Hospitalization	A0705	Battery Problem	10/11/2021	10/11/2021
									A070504	Premature Discharge of Battery	10/11/2021	10/11/2021
							F1905	Device Revision or Replacement	A0705	Battery Problem	10/11/2021	10/11/2021
									A070504	Premature Discharge of Battery	10/11/2021	10/11/2021
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem	10/11/2021	10/11/2021
									A070504	Premature Discharge of Battery	10/11/2021	10/11/2021
							F08	Hospitalization or Prolonged Hospitalization	A0705	Battery Problem	10/11/2021	10/11/2021
									A070504	Premature Discharge of Battery	10/11/2021	10/11/2021
							F1905	Device Revision or Replacement	A0705	Battery Problem	10/11/2021	10/11/2021
									A070504	Premature Discharge of Battery	10/11/2021	10/11/2021
					CT2985	Active	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem	10/11/2021	10/11/2021
									A070504	Premature Discharge of Battery	10/11/2021	10/11/2021
							F08	Hospitalization or Prolonged Hospitalization	A0705	Battery Problem	10/11/2021	10/11/2021
									A070504	Premature Discharge of Battery	10/11/2021	10/11/2021
							F1905	Device Revision or Replacement	A0705	Battery Problem	10/11/2021	10/11/2021
									A070504	Premature Discharge of Battery	10/11/2021	10/11/2021
					CT3001	Implantable metal	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem	10/11/2021	10/11/2021
									A070504	Premature Discharge of Battery	10/11/2021	10/11/2021
							F08	Hospitalization or Prolonged Hospitalization	A0705	Battery Problem	10/11/2021	10/11/2021
									A070504	Premature Discharge of Battery	10/11/2021	10/11/2021
							F1905	Device Revision or Replacement	A0705	Battery Problem	10/11/2021	10/11/2021
									A070504	Premature Discharge of Battery	10/11/2021	10/11/2021
					CT3002	Implantable synthetic polymer	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem	10/11/2021	10/11/2021
									A070504	Premature Discharge of Battery	10/11/2021	10/11/2021
							F08	Hospitalization or Prolonged Hospitalization	A0705	Battery Problem	10/11/2021	10/11/2021
									A070504	Premature Discharge of Battery	10/11/2021	10/11/2021
							F1905	Device Revision or Replacement	A0705	Battery Problem	10/11/2021	10/11/2021
									A070504	Premature Discharge of Battery	10/11/2021	10/11/2021
					CT3006	Implantable inorganic material	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem	10/11/2021	10/11/2021
									A070504	Premature Discharge of Battery	10/11/2021	10/11/2021
							F08	Hospitalization or Prolonged Hospitalization	A0705	Battery Problem	10/11/2021	10/11/2021
									A070504	Premature Discharge of Battery	10/11/2021	10/11/2021
							F1905	Device Revision or Replacement	A0705	Battery Problem	10/11/2021	10/11/2021
									A070504	Premature Discharge of Battery	10/11/2021	10/11/2021
					CT314	Electrically powered	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem	10/11/2021	10/11/2021
									A070504	Premature Discharge of Battery	10/11/2021	10/11/2021
							F08	Hospitalization or Prolonged Hospitalization	A0705	Battery Problem	10/11/2021	10/11/2021
									A070504	Premature Discharge of Battery	10/11/2021	10/11/2021
							F1905	Device Revision or Replacement	A0705	Battery Problem	10/11/2021	10/11/2021
									A070504	Premature Discharge of Battery	10/11/2021	10/11/2021
					CT321	Long-term surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem	10/11/2021	10/11/2021
									A070504	Premature Discharge of Battery	10/11/2021	10/11/2021
							F08	Hospitalization or Prolonged Hospitalization	A0705	Battery Problem	10/11/2021	10/11/2021
									A070504	Premature Discharge of Battery	10/11/2021	10/11/2021
							F1905	Device Revision or Replacement	A0705	Battery Problem	10/11/2021	10/11/2021
									A070504	Premature Discharge of Battery	10/11/2021	10/11/2021
					CT449	Generators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem	10/11/2021	10/11/2021
									A070504	Premature Discharge of Battery	10/11/2021	10/11/2021
							F08	Hospitalization or Prolonged Hospitalization	A0705	Battery Problem	10/11/2021	10/11/2021
									A070504	Premature Discharge of Battery	10/11/2021	10/11/2021
							F1905	Device Revision or Replacement	A0705	Battery Problem	10/11/2021	10/11/2021
									A070504	Premature Discharge of Battery	10/11/2021	10/11/2021
					CT981	Single use	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem	10/11/2021	10/11/2021
									A070504	Premature Discharge of Battery	10/11/2021	10/11/2021
							F08	Hospitalization or Prolonged Hospitalization	A0705	Battery Problem	10/11/2021	10/11/2021
									A070504	Premature Discharge of Battery	10/11/2021	10/11/2021
							F1905	Device Revision or Replacement	A0705	Battery Problem	10/11/2021	10/11/2021
									A070504	Premature Discharge of Battery	10/11/2021	10/11/2021
					CT983	Surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem	10/11/2021	10/11/2021
									A070504	Premature Discharge of Battery	10/11/2021	10/11/2021
							F08	Hospitalization or Prolonged Hospitalization	A0705	Battery Problem	10/11/2021	10/11/2021
									A070504	Premature Discharge of Battery	10/11/2021	10/11/2021
							F1905	Device Revision or Replacement	A0705	Battery Problem	10/11/2021	10/11/2021
									A070504	Premature Discharge of Battery	10/11/2021	10/11/2021
6878	2021	All other reportable incidents	47265	Dual-chamber implantable pacemaker, rate-responsive	CT141	Electrodes/Leads and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	01/04/2022	01/04/2022
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	01/04/2022	01/04/2022
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	01/04/2022	01/04/2022
					CT145	Cardiovascular devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	01/04/2022	01/04/2022
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	01/04/2022	01/04/2022
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	01/04/2022	01/04/2022
					CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	01/04/2022	01/04/2022
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	01/04/2022	01/04/2022
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	01/04/2022	01/04/2022
					CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	01/04/2022	01/04/2022
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	01/04/2022	01/04/2022

							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	01/04/2022	01/04/2022
					CT233	Surgical	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	01/04/2022	01/04/2022
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	01/04/2022	01/04/2022
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	01/04/2022	01/04/2022
					CT2406	Implantable	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	01/04/2022	01/04/2022
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	01/04/2022	01/04/2022
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	01/04/2022	01/04/2022
					CT2407	Active implantable	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	01/04/2022	01/04/2022
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	01/04/2022	01/04/2022
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	01/04/2022	01/04/2022
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	01/04/2022	01/04/2022
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	01/04/2022	01/04/2022
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	01/04/2022	01/04/2022
					CT2727	Active implant devices/ancillaries	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	01/04/2022	01/04/2022
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	01/04/2022	01/04/2022
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	01/04/2022	01/04/2022
					CT2750	Implantable cardiac electrical stimulators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	01/04/2022	01/04/2022
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	01/04/2022	01/04/2022
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	01/04/2022	01/04/2022
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	01/04/2022	01/04/2022
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	01/04/2022	01/04/2022
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	01/04/2022	01/04/2022
					CT2985	Active	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	01/04/2022	01/04/2022
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	01/04/2022	01/04/2022
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	01/04/2022	01/04/2022
					CT3001	Implantable metal	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	01/04/2022	01/04/2022
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	01/04/2022	01/04/2022
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	01/04/2022	01/04/2022
					CT3002	Implantable synthetic polymer	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	01/04/2022	01/04/2022
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	01/04/2022	01/04/2022
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	01/04/2022	01/04/2022
					CT3006	Implantable inorganic material	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	01/04/2022	01/04/2022
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	01/04/2022	01/04/2022
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	01/04/2022	01/04/2022
					CT314	Electrically powered	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	01/04/2022	01/04/2022
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	01/04/2022	01/04/2022
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	01/04/2022	01/04/2022
					CT321	Long-term surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	01/04/2022	01/04/2022
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	01/04/2022	01/04/2022
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	01/04/2022	01/04/2022
					CT449	Generators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	01/04/2022	01/04/2022
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	01/04/2022	01/04/2022
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	01/04/2022	01/04/2022
					CT981	Single use	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	01/04/2022	01/04/2022
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	01/04/2022	01/04/2022
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	01/04/2022	01/04/2022
					CT983	Surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	01/04/2022	01/04/2022
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	01/04/2022	01/04/2022
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	01/04/2022	01/04/2022
6879	2022	All other reportable incidents	17882	Physiologic-monitoring defibrillation system	CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A07	Electrical /Electronic Property Problem		
							F27	Problem Identified During Non-Clinical Procedure	A07	Electrical /Electronic Property Problem		
6880	2022	All other reportable incidents	35223	Endocardial/interventricular septal pacing lead	CT1684	Cardiac pacemakers and associated devices	E0204	Lethargy	A070101	Failure to Capture		
									A072201	High Impedance		
							E2312	Fatigue	A070101	Failure to Capture		
									A072201	High Impedance		
							E2403	No Clinical Signs, Symptoms or Conditions	A070101	Failure to Capture		
									A072201	High Impedance		
							F11	Non-Serious Injury/ Illness/ Impairment	A070101	Failure to Capture		
									A072201	High Impedance		
							F1905	Device Revision or Replacement	A070101	Failure to Capture		
									A072201	High Impedance		
6881	2022	All other reportable incidents	35852	Single-chamber implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	27/04/2022	27/04/2022
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	27/04/2022	27/04/2022
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	27/04/2022	27/04/2022
6882	2022	All other reportable incidents	47267	Single-chamber implantable pacemaker, rate-responsive	CT141	Electrodes/Leads and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem		
							F12	Serious Injury/ Illness/ Impairment	A0705	Battery Problem		
							F23	Unexpected Medical Intervention	A0705	Battery Problem		
					CT145	Cardiovascular devices	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem		
							F12	Serious Injury/ Illness/ Impairment	A0705	Battery Problem		
							F23	Unexpected Medical Intervention	A0705	Battery Problem		
					CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem		
							F12	Serious Injury/ Illness/ Impairment	A0705	Battery Problem		
							F23	Unexpected Medical Intervention	A0705	Battery Problem		
					CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem		
							F12	Serious Injury/ Illness/ Impairment	A0705	Battery Problem		
							F23	Unexpected Medical Intervention	A0705	Battery Problem		
					CT233	Surgical	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem		
							F12	Serious Injury/ Illness/ Impairment	A0705	Battery Problem		
							F23	Unexpected Medical Intervention	A0705	Battery Problem		
					CT2406	Implantable	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem		
							F12	Serious Injury/ Illness/ Impairment	A0705	Battery Problem		
							F23	Unexpected Medical Intervention	A0705	Battery Problem		
					CT2407	Active implantable	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem		
							F12	Serious Injury/ Illness/ Impairment	A0705	Battery Problem		
							F23	Unexpected Medical Intervention	A0705	Battery Problem		
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem		
							F12	Serious Injury/ Illness/ Impairment	A0705	Battery Problem		
							F23	Unexpected Medical Intervention	A0705	Battery Problem		

					CT2727	Active implant devices/ancillaries	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem		
							F12	Serious Injury/ Illness/ Impairment	A0705	Battery Problem		
							F23	Unexpected Medical Intervention	A0705	Battery Problem		
					CT2750	Implantable cardiac electrical stimulators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem		
							F12	Serious Injury/ Illness/ Impairment	A0705	Battery Problem		
							F23	Unexpected Medical Intervention	A0705	Battery Problem		
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem		
							F12	Serious Injury/ Illness/ Impairment	A0705	Battery Problem		
							F23	Unexpected Medical Intervention	A0705	Battery Problem		
					CT2985	Active	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem		
							F12	Serious Injury/ Illness/ Impairment	A0705	Battery Problem		
							F23	Unexpected Medical Intervention	A0705	Battery Problem		
					CT3001	Implantable metal	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem		
							F12	Serious Injury/ Illness/ Impairment	A0705	Battery Problem		
							F23	Unexpected Medical Intervention	A0705	Battery Problem		
					CT3002	Implantable synthetic polymer	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem		
							F12	Serious Injury/ Illness/ Impairment	A0705	Battery Problem		
							F23	Unexpected Medical Intervention	A0705	Battery Problem		
					CT3006	Implantable inorganic material	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem		
							F12	Serious Injury/ Illness/ Impairment	A0705	Battery Problem		
							F23	Unexpected Medical Intervention	A0705	Battery Problem		
					CT314	Electrically powered	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem		
							F12	Serious Injury/ Illness/ Impairment	A0705	Battery Problem		
							F23	Unexpected Medical Intervention	A0705	Battery Problem		
					CT321	Long-term surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem		
							F12	Serious Injury/ Illness/ Impairment	A0705	Battery Problem		
							F23	Unexpected Medical Intervention	A0705	Battery Problem		
					CT449	Generators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem		
							F12	Serious Injury/ Illness/ Impairment	A0705	Battery Problem		
							F23	Unexpected Medical Intervention	A0705	Battery Problem		
					CT981	Single use	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem		
							F12	Serious Injury/ Illness/ Impairment	A0705	Battery Problem		
							F23	Unexpected Medical Intervention	A0705	Battery Problem		
					CT983	Surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem		
							F12	Serious Injury/ Illness/ Impairment	A0705	Battery Problem		
							F23	Unexpected Medical Intervention	A0705	Battery Problem		
6883	2022	All other reportable incidents	47265	Dual-chamber implantable pacemaker, rate-responsive	CT141	Electrodes/Leads and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem	01/09/2022	01/09/2022
									A070504	Premature Discharge of Battery	01/09/2022	01/09/2022
							F08	Hospitalization or Prolonged Hospitalization	A0705	Battery Problem	01/09/2022	01/09/2022
									A070504	Premature Discharge of Battery	01/09/2022	01/09/2022
							F1903	Device Explantation	A0705	Battery Problem	01/09/2022	01/09/2022
									A070504	Premature Discharge of Battery	01/09/2022	01/09/2022
					CT145	Cardiovascular devices	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem	01/09/2022	01/09/2022
									A070504	Premature Discharge of Battery	01/09/2022	01/09/2022
							F08	Hospitalization or Prolonged Hospitalization	A0705	Battery Problem	01/09/2022	01/09/2022
									A070504	Premature Discharge of Battery	01/09/2022	01/09/2022
							F1903	Device Explantation	A0705	Battery Problem	01/09/2022	01/09/2022
									A070504	Premature Discharge of Battery	01/09/2022	01/09/2022
					CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem	01/09/2022	01/09/2022
									A070504	Premature Discharge of Battery	01/09/2022	01/09/2022
							F08	Hospitalization or Prolonged Hospitalization	A0705	Battery Problem	01/09/2022	01/09/2022
									A070504	Premature Discharge of Battery	01/09/2022	01/09/2022
							F1903	Device Explantation	A0705	Battery Problem	01/09/2022	01/09/2022
									A070504	Premature Discharge of Battery	01/09/2022	01/09/2022
					CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem	01/09/2022	01/09/2022
									A070504	Premature Discharge of Battery	01/09/2022	01/09/2022
							F08	Hospitalization or Prolonged Hospitalization	A0705	Battery Problem	01/09/2022	01/09/2022
									A070504	Premature Discharge of Battery	01/09/2022	01/09/2022
							F1903	Device Explantation	A0705	Battery Problem	01/09/2022	01/09/2022
									A070504	Premature Discharge of Battery	01/09/2022	01/09/2022
					CT233	Surgical	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem	01/09/2022	01/09/2022
									A070504	Premature Discharge of Battery	01/09/2022	01/09/2022
							F08	Hospitalization or Prolonged Hospitalization	A0705	Battery Problem	01/09/2022	01/09/2022
									A070504	Premature Discharge of Battery	01/09/2022	01/09/2022
							F1903	Device Explantation	A0705	Battery Problem	01/09/2022	01/09/2022
									A070504	Premature Discharge of Battery	01/09/2022	01/09/2022
					CT2406	Implantable	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem	01/09/2022	01/09/2022
									A070504	Premature Discharge of Battery	01/09/2022	01/09/2022
							F08	Hospitalization or Prolonged Hospitalization	A0705	Battery Problem	01/09/2022	01/09/2022
									A070504	Premature Discharge of Battery	01/09/2022	01/09/2022
							F1903	Device Explantation	A0705	Battery Problem	01/09/2022	01/09/2022
									A070504	Premature Discharge of Battery	01/09/2022	01/09/2022
					CT2407	Active implantable	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem	01/09/2022	01/09/2022
									A070504	Premature Discharge of Battery	01/09/2022	01/09/2022
							F08	Hospitalization or Prolonged Hospitalization	A0705	Battery Problem	01/09/2022	01/09/2022
									A070504	Premature Discharge of Battery	01/09/2022	01/09/2022
							F1903	Device Explantation	A0705	Battery Problem	01/09/2022	01/09/2022
									A070504	Premature Discharge of Battery	01/09/2022	01/09/2022
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem	01/09/2022	01/09/2022
									A070504	Premature Discharge of Battery	01/09/2022	01/09/2022
							F08	Hospitalization or Prolonged Hospitalization	A0705	Battery Problem	01/09/2022	01/09/2022
									A070504	Premature Discharge of Battery	01/09/2022	01/09/2022
							F1903	Device Explantation	A0705	Battery Problem	01/09/2022	01/09/2022
									A070504	Premature Discharge of Battery	01/09/2022	01/09/2022
					CT2727	Active implant devices/ancillaries	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem	01/09/2022	01/09/2022
									A070504	Premature Discharge of Battery	01/09/2022	01/09/2022
							F08	Hospitalization or Prolonged Hospitalization	A0705	Battery Problem	01/09/2022	01/09/2022
									A070504	Premature Discharge of Battery	01/09/2022	01/09/2022

							F1903	Device Explantation	A0705	Battery Problem	01/09/2022	01/09/2022
									A070504	Premature Discharge of Battery	01/09/2022	01/09/2022
					CT2750	Implantable cardiac electrical stimulators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem	01/09/2022	01/09/2022
									A070504	Premature Discharge of Battery	01/09/2022	01/09/2022
							F08	Hospitalization or Prolonged Hospitalization	A0705	Battery Problem	01/09/2022	01/09/2022
									A070504	Premature Discharge of Battery	01/09/2022	01/09/2022
							F1903	Device Explantation	A0705	Battery Problem	01/09/2022	01/09/2022
									A070504	Premature Discharge of Battery	01/09/2022	01/09/2022
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem	01/09/2022	01/09/2022
									A070504	Premature Discharge of Battery	01/09/2022	01/09/2022
							F08	Hospitalization or Prolonged Hospitalization	A0705	Battery Problem	01/09/2022	01/09/2022
									A070504	Premature Discharge of Battery	01/09/2022	01/09/2022
							F1903	Device Explantation	A0705	Battery Problem	01/09/2022	01/09/2022
									A070504	Premature Discharge of Battery	01/09/2022	01/09/2022
					CT2985	Active	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem	01/09/2022	01/09/2022
									A070504	Premature Discharge of Battery	01/09/2022	01/09/2022
							F08	Hospitalization or Prolonged Hospitalization	A0705	Battery Problem	01/09/2022	01/09/2022
									A070504	Premature Discharge of Battery	01/09/2022	01/09/2022
							F1903	Device Explantation	A0705	Battery Problem	01/09/2022	01/09/2022
									A070504	Premature Discharge of Battery	01/09/2022	01/09/2022
					CT3001	Implantable metal	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem	01/09/2022	01/09/2022
									A070504	Premature Discharge of Battery	01/09/2022	01/09/2022
							F08	Hospitalization or Prolonged Hospitalization	A0705	Battery Problem	01/09/2022	01/09/2022
									A070504	Premature Discharge of Battery	01/09/2022	01/09/2022
							F1903	Device Explantation	A0705	Battery Problem	01/09/2022	01/09/2022
									A070504	Premature Discharge of Battery	01/09/2022	01/09/2022
					CT3002	Implantable synthetic polymer	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem	01/09/2022	01/09/2022
									A070504	Premature Discharge of Battery	01/09/2022	01/09/2022
							F08	Hospitalization or Prolonged Hospitalization	A0705	Battery Problem	01/09/2022	01/09/2022
									A070504	Premature Discharge of Battery	01/09/2022	01/09/2022
							F1903	Device Explantation	A0705	Battery Problem	01/09/2022	01/09/2022
									A070504	Premature Discharge of Battery	01/09/2022	01/09/2022
					CT3006	Implantable inorganic material	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem	01/09/2022	01/09/2022
									A070504	Premature Discharge of Battery	01/09/2022	01/09/2022
							F08	Hospitalization or Prolonged Hospitalization	A0705	Battery Problem	01/09/2022	01/09/2022
									A070504	Premature Discharge of Battery	01/09/2022	01/09/2022
							F1903	Device Explantation	A0705	Battery Problem	01/09/2022	01/09/2022
									A070504	Premature Discharge of Battery	01/09/2022	01/09/2022
					CT314	Electrically powered	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem	01/09/2022	01/09/2022
									A070504	Premature Discharge of Battery	01/09/2022	01/09/2022
							F08	Hospitalization or Prolonged Hospitalization	A0705	Battery Problem	01/09/2022	01/09/2022
									A070504	Premature Discharge of Battery	01/09/2022	01/09/2022
							F1903	Device Explantation	A0705	Battery Problem	01/09/2022	01/09/2022
									A070504	Premature Discharge of Battery	01/09/2022	01/09/2022
					CT321	Long-term surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem	01/09/2022	01/09/2022
									A070504	Premature Discharge of Battery	01/09/2022	01/09/2022
							F08	Hospitalization or Prolonged Hospitalization	A0705	Battery Problem	01/09/2022	01/09/2022
									A070504	Premature Discharge of Battery	01/09/2022	01/09/2022
							F1903	Device Explantation	A0705	Battery Problem	01/09/2022	01/09/2022
									A070504	Premature Discharge of Battery	01/09/2022	01/09/2022
					CT449	Generators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem	01/09/2022	01/09/2022
									A070504	Premature Discharge of Battery	01/09/2022	01/09/2022
							F08	Hospitalization or Prolonged Hospitalization	A0705	Battery Problem	01/09/2022	01/09/2022
									A070504	Premature Discharge of Battery	01/09/2022	01/09/2022
							F1903	Device Explantation	A0705	Battery Problem	01/09/2022	01/09/2022
									A070504	Premature Discharge of Battery	01/09/2022	01/09/2022
					CT981	Single use	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem	01/09/2022	01/09/2022
									A070504	Premature Discharge of Battery	01/09/2022	01/09/2022
							F08	Hospitalization or Prolonged Hospitalization	A0705	Battery Problem	01/09/2022	01/09/2022
									A070504	Premature Discharge of Battery	01/09/2022	01/09/2022
							F1903	Device Explantation	A0705	Battery Problem	01/09/2022	01/09/2022
									A070504	Premature Discharge of Battery	01/09/2022	01/09/2022
					CT983	Surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem	01/09/2022	01/09/2022
									A070504	Premature Discharge of Battery	01/09/2022	01/09/2022
							F08	Hospitalization or Prolonged Hospitalization	A0705	Battery Problem	01/09/2022	01/09/2022
									A070504	Premature Discharge of Battery	01/09/2022	01/09/2022
							F1903	Device Explantation	A0705	Battery Problem	01/09/2022	01/09/2022
									A070504	Premature Discharge of Battery	01/09/2022	01/09/2022
6884	2022	All other reportable incidents	35852	Single-chamber implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E2104	Electric Shock	A070910	Undersensing		
									A071302	Inappropriate/Inadequate Shock/Stimulation		
									A090801	Signal Artifact/Noise		
									A1301	Failure to Read Input Signal		
							E2403	No Clinical Signs, Symptoms or Conditions	A070910	Undersensing		
									A071302	Inappropriate/Inadequate Shock/Stimulation		
									A090801	Signal Artifact/Noise		
									A1301	Failure to Read Input Signal		
							F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A070910	Undersensing		
									A071302	Inappropriate/Inadequate Shock/Stimulation		
									A090801	Signal Artifact/Noise		
									A1301	Failure to Read Input Signal		
							F2203	Imaging Required	A070910	Undersensing		
									A071302	Inappropriate/Inadequate Shock/Stimulation		
									A090801	Signal Artifact/Noise		
									A1301	Failure to Read Input Signal		
6885	2022	All other reportable incidents	47106	Drug-eluting endocardial pacing lead	CT1684	Cardiac pacemakers and associated devices	E2401	Insufficient Information	A070909	Oversensing		
							E2403	No Clinical Signs, Symptoms or Conditions	A070909	Oversensing		
							F01	Change in Therapeutic Response	A070909	Oversensing		
6886	2022	All other reportable incidents	17882	Physiologic-monitoring defibrillation system	CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070903	Failure to Analyze Signal		

							F27	Problem Identified During Non-Clinical Procedure	A070903	Failure to Analyze Signal		
6887	2022	All other reportable incidents	37265	Dual-chamber implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	12/05/2022	12/05/2022
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	12/05/2022	12/05/2022
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	12/05/2022	12/05/2022
6888	2021	All other reportable incidents	47265	Dual-chamber implantable pacemaker, rate-responsive	CT141	Electrodes/Leads and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
					CT145	Cardiovascular devices	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
					CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
					CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
					CT233	Surgical	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
					CT2406	Implantable	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
					CT2407	Active implantable	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
					CT2727	Active implant devices/ancillaries	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
					CT2750	Implantable cardiac electrical stimulators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
					CT2985	Active	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
					CT3001	Implantable metal	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
					CT3002	Implantable synthetic polymer	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
					CT3006	Implantable inorganic material	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
					CT314	Electrically powered	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
					CT321	Long-term surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
					CT449	Generators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
					CT981	Single use	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
					CT983	Surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
6889	2022	All other reportable incidents	47265	Dual-chamber implantable pacemaker, rate-responsive	CT141	Electrodes/Leads and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem	06/06/2022	06/06/2022
									A070504	Premature Discharge of Battery	06/06/2022	06/06/2022
									A0722	Impedance Problem	06/06/2022	06/06/2022

								A0908	Incorrect, Inadequate or Imprecise Result or Readings	06/06/2022	06/06/2022
						F08	Hospitalization or Prolonged Hospitalization	A0705	Battery Problem	06/06/2022	06/06/2022
								A070504	Premature Discharge of Battery	06/06/2022	06/06/2022
								A0722	Impedance Problem	06/06/2022	06/06/2022
								A0908	Incorrect, Inadequate or Imprecise Result or Readings	06/06/2022	06/06/2022
						F1905	Device Revision or Replacement	A0705	Battery Problem	06/06/2022	06/06/2022
								A070504	Premature Discharge of Battery	06/06/2022	06/06/2022
								A0722	Impedance Problem	06/06/2022	06/06/2022
								A0908	Incorrect, Inadequate or Imprecise Result or Readings	06/06/2022	06/06/2022
					CT145	Cardiovascular devices	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem	06/06/2022	06/06/2022
									A070504	Premature Discharge of Battery	06/06/2022	06/06/2022
									A0722	Impedance Problem	06/06/2022	06/06/2022
									A0908	Incorrect, Inadequate or Imprecise Result or Readings	06/06/2022	06/06/2022
							F08	Hospitalization or Prolonged Hospitalization	A0705	Battery Problem	06/06/2022	06/06/2022
									A070504	Premature Discharge of Battery	06/06/2022	06/06/2022
									A0722	Impedance Problem	06/06/2022	06/06/2022
									A0908	Incorrect, Inadequate or Imprecise Result or Readings	06/06/2022	06/06/2022
							F1905	Device Revision or Replacement	A0705	Battery Problem	06/06/2022	06/06/2022
									A070504	Premature Discharge of Battery	06/06/2022	06/06/2022
									A0722	Impedance Problem	06/06/2022	06/06/2022
									A0908	Incorrect, Inadequate or Imprecise Result or Readings	06/06/2022	06/06/2022
					CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem	06/06/2022	06/06/2022
									A070504	Premature Discharge of Battery	06/06/2022	06/06/2022
									A0722	Impedance Problem	06/06/2022	06/06/2022
									A0908	Incorrect, Inadequate or Imprecise Result or Readings	06/06/2022	06/06/2022
							F08	Hospitalization or Prolonged Hospitalization	A0705	Battery Problem	06/06/2022	06/06/2022
									A070504	Premature Discharge of Battery	06/06/2022	06/06/2022
									A0722	Impedance Problem	06/06/2022	06/06/2022
									A0908	Incorrect, Inadequate or Imprecise Result or Readings	06/06/2022	06/06/2022
							F1905	Device Revision or Replacement	A0705	Battery Problem	06/06/2022	06/06/2022
									A070504	Premature Discharge of Battery	06/06/2022	06/06/2022
									A0722	Impedance Problem	06/06/2022	06/06/2022
									A0908	Incorrect, Inadequate or Imprecise Result or Readings	06/06/2022	06/06/2022
					CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem	06/06/2022	06/06/2022
									A070504	Premature Discharge of Battery	06/06/2022	06/06/2022
									A0722	Impedance Problem	06/06/2022	06/06/2022
									A0908	Incorrect, Inadequate or Imprecise Result or Readings	06/06/2022	06/06/2022
							F08	Hospitalization or Prolonged Hospitalization	A0705	Battery Problem	06/06/2022	06/06/2022
									A070504	Premature Discharge of Battery	06/06/2022	06/06/2022
									A0722	Impedance Problem	06/06/2022	06/06/2022
									A0908	Incorrect, Inadequate or Imprecise Result or Readings	06/06/2022	06/06/2022
							F1905	Device Revision or Replacement	A0705	Battery Problem	06/06/2022	06/06/2022
									A070504	Premature Discharge of Battery	06/06/2022	06/06/2022
									A0722	Impedance Problem	06/06/2022	06/06/2022
									A0908	Incorrect, Inadequate or Imprecise Result or Readings	06/06/2022	06/06/2022
					CT233	Surgical	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem	06/06/2022	06/06/2022
									A070504	Premature Discharge of Battery	06/06/2022	06/06/2022
									A0722	Impedance Problem	06/06/2022	06/06/2022
									A0908	Incorrect, Inadequate or Imprecise Result or Readings	06/06/2022	06/06/2022
							F08	Hospitalization or Prolonged Hospitalization	A0705	Battery Problem	06/06/2022	06/06/2022
									A070504	Premature Discharge of Battery	06/06/2022	06/06/2022
									A0722	Impedance Problem	06/06/2022	06/06/2022
									A0908	Incorrect, Inadequate or Imprecise Result or Readings	06/06/2022	06/06/2022
							F1905	Device Revision or Replacement	A0705	Battery Problem	06/06/2022	06/06/2022
									A070504	Premature Discharge of Battery	06/06/2022	06/06/2022
									A0722	Impedance Problem	06/06/2022	06/06/2022
									A0908	Incorrect, Inadequate or Imprecise Result or Readings	06/06/2022	06/06/2022
					CT2406	Implantable	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem	06/06/2022	06/06/2022
									A070504	Premature Discharge of Battery	06/06/2022	06/06/2022
									A0722	Impedance Problem	06/06/2022	06/06/2022
									A0908	Incorrect, Inadequate or Imprecise Result or Readings	06/06/2022	06/06/2022
							F08	Hospitalization or Prolonged Hospitalization	A0705	Battery Problem	06/06/2022	06/06/2022
									A070504	Premature Discharge of Battery	06/06/2022	06/06/2022
									A0722	Impedance Problem	06/06/2022	06/06/2022
									A0908	Incorrect, Inadequate or Imprecise Result or Readings	06/06/2022	06/06/2022
							F1905	Device Revision or Replacement	A0705	Battery Problem	06/06/2022	06/06/2022
									A070504	Premature Discharge of Battery	06/06/2022	06/06/2022
									A0722	Impedance Problem	06/06/2022	06/06/2022
									A0908	Incorrect, Inadequate or Imprecise Result or Readings	06/06/2022	06/06/2022
					CT2407	Active implantable	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem	06/06/2022	06/06/2022
									A070504	Premature Discharge of Battery	06/06/2022	06/06/2022
									A0722	Impedance Problem	06/06/2022	06/06/2022
									A0908	Incorrect, Inadequate or Imprecise Result or Readings	06/06/2022	06/06/2022

							F08	Hospitalization or Prolonged Hospitalization	A0705	Battery Problem	06/06/2022	06/06/2022
									A070504	Premature Discharge of Battery	06/06/2022	06/06/2022
									A0722	Impedance Problem	06/06/2022	06/06/2022
									A0908	Incorrect, Inadequate or Imprecise Result or Readings	06/06/2022	06/06/2022
							F1905	Device Revision or Replacement	A0705	Battery Problem	06/06/2022	06/06/2022
									A070504	Premature Discharge of Battery	06/06/2022	06/06/2022
									A0722	Impedance Problem	06/06/2022	06/06/2022
									A0908	Incorrect, Inadequate or Imprecise Result or Readings	06/06/2022	06/06/2022
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem	06/06/2022	06/06/2022
									A070504	Premature Discharge of Battery	06/06/2022	06/06/2022
									A0722	Impedance Problem	06/06/2022	06/06/2022
									A0908	Incorrect, Inadequate or Imprecise Result or Readings	06/06/2022	06/06/2022
							F08	Hospitalization or Prolonged Hospitalization	A0705	Battery Problem	06/06/2022	06/06/2022
									A070504	Premature Discharge of Battery	06/06/2022	06/06/2022
									A0722	Impedance Problem	06/06/2022	06/06/2022
									A0908	Incorrect, Inadequate or Imprecise Result or Readings	06/06/2022	06/06/2022
							F1905	Device Revision or Replacement	A0705	Battery Problem	06/06/2022	06/06/2022
									A070504	Premature Discharge of Battery	06/06/2022	06/06/2022
									A0722	Impedance Problem	06/06/2022	06/06/2022
									A0908	Incorrect, Inadequate or Imprecise Result or Readings	06/06/2022	06/06/2022
					CT2727	Active implant devices/ancillaries	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem	06/06/2022	06/06/2022
									A070504	Premature Discharge of Battery	06/06/2022	06/06/2022
									A0722	Impedance Problem	06/06/2022	06/06/2022
									A0908	Incorrect, Inadequate or Imprecise Result or Readings	06/06/2022	06/06/2022
							F08	Hospitalization or Prolonged Hospitalization	A0705	Battery Problem	06/06/2022	06/06/2022
									A070504	Premature Discharge of Battery	06/06/2022	06/06/2022
									A0722	Impedance Problem	06/06/2022	06/06/2022
									A0908	Incorrect, Inadequate or Imprecise Result or Readings	06/06/2022	06/06/2022
							F1905	Device Revision or Replacement	A0705	Battery Problem	06/06/2022	06/06/2022
									A070504	Premature Discharge of Battery	06/06/2022	06/06/2022
									A0722	Impedance Problem	06/06/2022	06/06/2022
									A0908	Incorrect, Inadequate or Imprecise Result or Readings	06/06/2022	06/06/2022
					CT2750	Implantable cardiac electrical stimulators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem	06/06/2022	06/06/2022
									A070504	Premature Discharge of Battery	06/06/2022	06/06/2022
									A0722	Impedance Problem	06/06/2022	06/06/2022
									A0908	Incorrect, Inadequate or Imprecise Result or Readings	06/06/2022	06/06/2022
							F08	Hospitalization or Prolonged Hospitalization	A0705	Battery Problem	06/06/2022	06/06/2022
									A070504	Premature Discharge of Battery	06/06/2022	06/06/2022
									A0722	Impedance Problem	06/06/2022	06/06/2022
									A0908	Incorrect, Inadequate or Imprecise Result or Readings	06/06/2022	06/06/2022
							F1905	Device Revision or Replacement	A0705	Battery Problem	06/06/2022	06/06/2022
									A070504	Premature Discharge of Battery	06/06/2022	06/06/2022
									A0722	Impedance Problem	06/06/2022	06/06/2022
									A0908	Incorrect, Inadequate or Imprecise Result or Readings	06/06/2022	06/06/2022
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem	06/06/2022	06/06/2022
									A070504	Premature Discharge of Battery	06/06/2022	06/06/2022
									A0722	Impedance Problem	06/06/2022	06/06/2022
									A0908	Incorrect, Inadequate or Imprecise Result or Readings	06/06/2022	06/06/2022
							F08	Hospitalization or Prolonged Hospitalization	A0705	Battery Problem	06/06/2022	06/06/2022
									A070504	Premature Discharge of Battery	06/06/2022	06/06/2022
									A0722	Impedance Problem	06/06/2022	06/06/2022
									A0908	Incorrect, Inadequate or Imprecise Result or Readings	06/06/2022	06/06/2022
							F1905	Device Revision or Replacement	A0705	Battery Problem	06/06/2022	06/06/2022
									A070504	Premature Discharge of Battery	06/06/2022	06/06/2022
									A0722	Impedance Problem	06/06/2022	06/06/2022
									A0908	Incorrect, Inadequate or Imprecise Result or Readings	06/06/2022	06/06/2022
					CT2985	Active	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem	06/06/2022	06/06/2022
									A070504	Premature Discharge of Battery	06/06/2022	06/06/2022
									A0722	Impedance Problem	06/06/2022	06/06/2022
									A0908	Incorrect, Inadequate or Imprecise Result or Readings	06/06/2022	06/06/2022
							F08	Hospitalization or Prolonged Hospitalization	A0705	Battery Problem	06/06/2022	06/06/2022
									A070504	Premature Discharge of Battery	06/06/2022	06/06/2022
									A0722	Impedance Problem	06/06/2022	06/06/2022
									A0908	Incorrect, Inadequate or Imprecise Result or Readings	06/06/2022	06/06/2022
							F1905	Device Revision or Replacement	A0705	Battery Problem	06/06/2022	06/06/2022
									A070504	Premature Discharge of Battery	06/06/2022	06/06/2022
									A0722	Impedance Problem	06/06/2022	06/06/2022
									A0908	Incorrect, Inadequate or Imprecise Result or Readings	06/06/2022	06/06/2022
					CT3001	Implantable metal	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem	06/06/2022	06/06/2022
									A070504	Premature Discharge of Battery	06/06/2022	06/06/2022
									A0722	Impedance Problem	06/06/2022	06/06/2022
									A0908	Incorrect, Inadequate or Imprecise Result or Readings	06/06/2022	06/06/2022
							F08	Hospitalization or Prolonged Hospitalization	A0705	Battery Problem	06/06/2022	06/06/2022
									A070504	Premature Discharge of Battery	06/06/2022	06/06/2022

								A0722	Impedance Problem	06/06/2022	06/06/2022
								A0908	Incorrect, Inadequate or Imprecise Result or Readings	06/06/2022	06/06/2022
						F1905	Device Revision or Replacement	A0705	Battery Problem	06/06/2022	06/06/2022
								A070504	Premature Discharge of Battery	06/06/2022	06/06/2022
								A0722	Impedance Problem	06/06/2022	06/06/2022
								A0908	Incorrect, Inadequate or Imprecise Result or Readings	06/06/2022	06/06/2022
					CT3002	Implantable synthetic polymer	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem	06/06/2022	06/06/2022
									A070504	Premature Discharge of Battery	06/06/2022	06/06/2022
									A0722	Impedance Problem	06/06/2022	06/06/2022
									A0908	Incorrect, Inadequate or Imprecise Result or Readings	06/06/2022	06/06/2022
							F08	Hospitalization or Prolonged Hospitalization	A0705	Battery Problem	06/06/2022	06/06/2022
									A070504	Premature Discharge of Battery	06/06/2022	06/06/2022
									A0722	Impedance Problem	06/06/2022	06/06/2022
									A0908	Incorrect, Inadequate or Imprecise Result or Readings	06/06/2022	06/06/2022
							F1905	Device Revision or Replacement	A0705	Battery Problem	06/06/2022	06/06/2022
									A070504	Premature Discharge of Battery	06/06/2022	06/06/2022
									A0722	Impedance Problem	06/06/2022	06/06/2022
									A0908	Incorrect, Inadequate or Imprecise Result or Readings	06/06/2022	06/06/2022
					CT3006	Implantable inorganic material	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem	06/06/2022	06/06/2022
									A070504	Premature Discharge of Battery	06/06/2022	06/06/2022
									A0722	Impedance Problem	06/06/2022	06/06/2022
									A0908	Incorrect, Inadequate or Imprecise Result or Readings	06/06/2022	06/06/2022
							F08	Hospitalization or Prolonged Hospitalization	A0705	Battery Problem	06/06/2022	06/06/2022
									A070504	Premature Discharge of Battery	06/06/2022	06/06/2022
									A0722	Impedance Problem	06/06/2022	06/06/2022
									A0908	Incorrect, Inadequate or Imprecise Result or Readings	06/06/2022	06/06/2022
							F1905	Device Revision or Replacement	A0705	Battery Problem	06/06/2022	06/06/2022
									A070504	Premature Discharge of Battery	06/06/2022	06/06/2022
									A0722	Impedance Problem	06/06/2022	06/06/2022
									A0908	Incorrect, Inadequate or Imprecise Result or Readings	06/06/2022	06/06/2022
					CT314	Electrically powered	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem	06/06/2022	06/06/2022
									A070504	Premature Discharge of Battery	06/06/2022	06/06/2022
									A0722	Impedance Problem	06/06/2022	06/06/2022
									A0908	Incorrect, Inadequate or Imprecise Result or Readings	06/06/2022	06/06/2022
							F08	Hospitalization or Prolonged Hospitalization	A0705	Battery Problem	06/06/2022	06/06/2022
									A070504	Premature Discharge of Battery	06/06/2022	06/06/2022
									A0722	Impedance Problem	06/06/2022	06/06/2022
									A0908	Incorrect, Inadequate or Imprecise Result or Readings	06/06/2022	06/06/2022
							F1905	Device Revision or Replacement	A0705	Battery Problem	06/06/2022	06/06/2022
									A070504	Premature Discharge of Battery	06/06/2022	06/06/2022
									A0722	Impedance Problem	06/06/2022	06/06/2022
									A0908	Incorrect, Inadequate or Imprecise Result or Readings	06/06/2022	06/06/2022
					CT321	Long-term surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem	06/06/2022	06/06/2022
									A070504	Premature Discharge of Battery	06/06/2022	06/06/2022
									A0722	Impedance Problem	06/06/2022	06/06/2022
									A0908	Incorrect, Inadequate or Imprecise Result or Readings	06/06/2022	06/06/2022
							F08	Hospitalization or Prolonged Hospitalization	A0705	Battery Problem	06/06/2022	06/06/2022
									A070504	Premature Discharge of Battery	06/06/2022	06/06/2022
									A0722	Impedance Problem	06/06/2022	06/06/2022
									A0908	Incorrect, Inadequate or Imprecise Result or Readings	06/06/2022	06/06/2022
							F1905	Device Revision or Replacement	A0705	Battery Problem	06/06/2022	06/06/2022
									A070504	Premature Discharge of Battery	06/06/2022	06/06/2022
									A0722	Impedance Problem	06/06/2022	06/06/2022
									A0908	Incorrect, Inadequate or Imprecise Result or Readings	06/06/2022	06/06/2022
					CT449	Generators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem	06/06/2022	06/06/2022
									A070504	Premature Discharge of Battery	06/06/2022	06/06/2022
									A0722	Impedance Problem	06/06/2022	06/06/2022
									A0908	Incorrect, Inadequate or Imprecise Result or Readings	06/06/2022	06/06/2022
							F08	Hospitalization or Prolonged Hospitalization	A0705	Battery Problem	06/06/2022	06/06/2022
									A070504	Premature Discharge of Battery	06/06/2022	06/06/2022
									A0722	Impedance Problem	06/06/2022	06/06/2022
									A0908	Incorrect, Inadequate or Imprecise Result or Readings	06/06/2022	06/06/2022
							F1905	Device Revision or Replacement	A0705	Battery Problem	06/06/2022	06/06/2022
									A070504	Premature Discharge of Battery	06/06/2022	06/06/2022
									A0722	Impedance Problem	06/06/2022	06/06/2022
									A0908	Incorrect, Inadequate or Imprecise Result or Readings	06/06/2022	06/06/2022
					CT981	Single use	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem	06/06/2022	06/06/2022
									A070504	Premature Discharge of Battery	06/06/2022	06/06/2022
									A0722	Impedance Problem	06/06/2022	06/06/2022
									A0908	Incorrect, Inadequate or Imprecise Result or Readings	06/06/2022	06/06/2022
							F08	Hospitalization or Prolonged Hospitalization	A0705	Battery Problem	06/06/2022	06/06/2022
									A070504	Premature Discharge of Battery	06/06/2022	06/06/2022
									A0722	Impedance Problem	06/06/2022	06/06/2022

								A0908	Incorrect, Inadequate or Imprecise Result or Readings	06/06/2022	06/06/2022
						F1905	Device Revision or Replacement	A0705	Battery Problem	06/06/2022	06/06/2022
								A070504	Premature Discharge of Battery	06/06/2022	06/06/2022
								A0722	Impedance Problem	06/06/2022	06/06/2022
								A0908	Incorrect, Inadequate or Imprecise Result or Readings	06/06/2022	06/06/2022
				CT983	Surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem	06/06/2022	06/06/2022
								A070504	Premature Discharge of Battery	06/06/2022	06/06/2022
								A0722	Impedance Problem	06/06/2022	06/06/2022
								A0908	Incorrect, Inadequate or Imprecise Result or Readings	06/06/2022	06/06/2022
						F08	Hospitalization or Prolonged Hospitalization	A0705	Battery Problem	06/06/2022	06/06/2022
								A070504	Premature Discharge of Battery	06/06/2022	06/06/2022
								A0722	Impedance Problem	06/06/2022	06/06/2022
								A0908	Incorrect, Inadequate or Imprecise Result or Readings	06/06/2022	06/06/2022
						F1905	Device Revision or Replacement	A0705	Battery Problem	06/06/2022	06/06/2022
								A070504	Premature Discharge of Battery	06/06/2022	06/06/2022
								A0722	Impedance Problem	06/06/2022	06/06/2022
								A0908	Incorrect, Inadequate or Imprecise Result or Readings	06/06/2022	06/06/2022
6890	2021	All other reportable incidents	35852	Single-chamber implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	23/12/2021	23/12/2021
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	23/12/2021	23/12/2021
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	23/12/2021	23/12/2021
6891	2022	All other reportable incidents	47270	Cardiac resynchronization therapy implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A072201	High Impedance		
							F26	No Health Consequences or Impact	A072201	High Impedance		
6892	2022	All other reportable incidents	36241	Pacemaker/defibrillator lead	CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A072201	High Impedance		
							F26	No Health Consequences or Impact	A072201	High Impedance		
6893	2021	All other reportable incidents	47265	Dual-chamber implantable pacemaker, rate-responsive	CT141	Electrodes/Leads and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A0203	Defective Device		
							F26	No Health Consequences or Impact	A0203	Defective Device		
					CT145	Cardiovascular devices	E2403	No Clinical Signs, Symptoms or Conditions	A0203	Defective Device		
							F26	No Health Consequences or Impact	A0203	Defective Device		
					CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A0203	Defective Device		
							F26	No Health Consequences or Impact	A0203	Defective Device		
					CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A0203	Defective Device		
							F26	No Health Consequences or Impact	A0203	Defective Device		
					CT233	Surgical	E2403	No Clinical Signs, Symptoms or Conditions	A0203	Defective Device		
							F26	No Health Consequences or Impact	A0203	Defective Device		
					CT2406	Implantable	E2403	No Clinical Signs, Symptoms or Conditions	A0203	Defective Device		
							F26	No Health Consequences or Impact	A0203	Defective Device		
					CT2407	Active implantable	E2403	No Clinical Signs, Symptoms or Conditions	A0203	Defective Device		
							F26	No Health Consequences or Impact	A0203	Defective Device		
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A0203	Defective Device		
							F26	No Health Consequences or Impact	A0203	Defective Device		
					CT2727	Active implant devices/ancillaries	E2403	No Clinical Signs, Symptoms or Conditions	A0203	Defective Device		
							F26	No Health Consequences or Impact	A0203	Defective Device		
					CT2750	Implantable cardiac electrical stimulators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A0203	Defective Device		
							F26	No Health Consequences or Impact	A0203	Defective Device		
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A0203	Defective Device		
							F26	No Health Consequences or Impact	A0203	Defective Device		
					CT2985	Active	E2403	No Clinical Signs, Symptoms or Conditions	A0203	Defective Device		
							F26	No Health Consequences or Impact	A0203	Defective Device		
					CT3001	Implantable metal	E2403	No Clinical Signs, Symptoms or Conditions	A0203	Defective Device		
							F26	No Health Consequences or Impact	A0203	Defective Device		
					CT3002	Implantable synthetic polymer	E2403	No Clinical Signs, Symptoms or Conditions	A0203	Defective Device		
							F26	No Health Consequences or Impact	A0203	Defective Device		
					CT3006	Implantable inorganic material	E2403	No Clinical Signs, Symptoms or Conditions	A0203	Defective Device		
							F26	No Health Consequences or Impact	A0203	Defective Device		
					CT314	Electrically powered	E2403	No Clinical Signs, Symptoms or Conditions	A0203	Defective Device		
							F26	No Health Consequences or Impact	A0203	Defective Device		
					CT321	Long-term surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A0203	Defective Device		
							F26	No Health Consequences or Impact	A0203	Defective Device		
					CT449	Generators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A0203	Defective Device		
							F26	No Health Consequences or Impact	A0203	Defective Device		
					CT981	Single use	E2403	No Clinical Signs, Symptoms or Conditions	A0203	Defective Device		
							F26	No Health Consequences or Impact	A0203	Defective Device		
					CT983	Surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A0203	Defective Device		
							F26	No Health Consequences or Impact	A0203	Defective Device		
6894	2021	All other reportable incidents	35852	Single-chamber implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E2104	Electric Shock	A070909	Oversensing		
									A071302	Inappropriate/Inadequate Shock/Stimulation		
									A1301	Failure to Read Input Signal		
							E2403	No Clinical Signs, Symptoms or Conditions	A070909	Oversensing		
									A071302	Inappropriate/Inadequate Shock/Stimulation		
									A1301	Failure to Read Input Signal		
							F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A070909	Oversensing		
									A071302	Inappropriate/Inadequate Shock/Stimulation		
									A1301	Failure to Read Input Signal		
6895	2022	All other reportable incidents	17882	Physiologic-monitoring defibrillation system	CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070912	Incorrect Interpretation of Signal		
							F27	Problem Identified During Non-Clinical Procedure	A070912	Incorrect Interpretation of Signal		
6896	2022	All other reportable incidents	36241	Pacemaker/defibrillator lead	CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A072201	High Impedance		
							F26	No Health Consequences or Impact	A072201	High Impedance		
6897	2022	All other reportable incidents	47265	Dual-chamber implantable pacemaker, rate-responsive	CT141	Electrodes/Leads and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
					CT145	Cardiovascular devices	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A0705	Battery Problem		

									A070504	Premature Discharge of Battery			
					CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem			
									A070504	Premature Discharge of Battery			
							F26	No Health Consequences or Impact	A0705	Battery Problem			
									A070504	Premature Discharge of Battery			
					CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem			
									A070504	Premature Discharge of Battery			
							F26	No Health Consequences or Impact	A0705	Battery Problem			
									A070504	Premature Discharge of Battery			
					CT233	Surgical	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem			
									A070504	Premature Discharge of Battery			
							F26	No Health Consequences or Impact	A0705	Battery Problem			
									A070504	Premature Discharge of Battery			
					CT2406	Implantable	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem			
									A070504	Premature Discharge of Battery			
							F26	No Health Consequences or Impact	A0705	Battery Problem			
									A070504	Premature Discharge of Battery			
					CT2407	Active implantable	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem			
									A070504	Premature Discharge of Battery			
							F26	No Health Consequences or Impact	A0705	Battery Problem			
									A070504	Premature Discharge of Battery			
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem			
									A070504	Premature Discharge of Battery			
							F26	No Health Consequences or Impact	A0705	Battery Problem			
									A070504	Premature Discharge of Battery			
					CT2727	Active implant devices/ancillaries	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem			
									A070504	Premature Discharge of Battery			
							F26	No Health Consequences or Impact	A0705	Battery Problem			
									A070504	Premature Discharge of Battery			
					CT2750	Implantable cardiac electrical stimulators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem			
									A070504	Premature Discharge of Battery			
							F26	No Health Consequences or Impact	A0705	Battery Problem			
									A070504	Premature Discharge of Battery			
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem			
									A070504	Premature Discharge of Battery			
							F26	No Health Consequences or Impact	A0705	Battery Problem			
									A070504	Premature Discharge of Battery			
					CT2985	Active	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem			
									A070504	Premature Discharge of Battery			
							F26	No Health Consequences or Impact	A0705	Battery Problem			
									A070504	Premature Discharge of Battery			
					CT3001	Implantable metal	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem			
									A070504	Premature Discharge of Battery			
							F26	No Health Consequences or Impact	A0705	Battery Problem			
									A070504	Premature Discharge of Battery			
					CT3002	Implantable synthetic polymer	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem			
									A070504	Premature Discharge of Battery			
							F26	No Health Consequences or Impact	A0705	Battery Problem			
									A070504	Premature Discharge of Battery			
					CT3006	Implantable inorganic material	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem			
									A070504	Premature Discharge of Battery			
							F26	No Health Consequences or Impact	A0705	Battery Problem			
									A070504	Premature Discharge of Battery			
					CT314	Electrically powered	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem			
									A070504	Premature Discharge of Battery			
							F26	No Health Consequences or Impact	A0705	Battery Problem			
									A070504	Premature Discharge of Battery			
					CT321	Long-term surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem			
									A070504	Premature Discharge of Battery			
							F26	No Health Consequences or Impact	A0705	Battery Problem			
									A070504	Premature Discharge of Battery			
					CT449	Generators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem			
									A070504	Premature Discharge of Battery			
							F26	No Health Consequences or Impact	A0705	Battery Problem			
									A070504	Premature Discharge of Battery			
					CT981	Single use	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem			
									A070504	Premature Discharge of Battery			
							F26	No Health Consequences or Impact	A0705	Battery Problem			
									A070504	Premature Discharge of Battery			
					CT983	Surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem			
									A070504	Premature Discharge of Battery			
							F26	No Health Consequences or Impact	A0705	Battery Problem			
									A070504	Premature Discharge of Battery			
6898	2021	All other reportable incidents	47270	Cardiac resynchronization therapy implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	03/11/2021	03/11/2021	
							F1901	Additional Surgery	A070504	Premature Discharge of Battery	03/11/2021	03/11/2021	
6899	2021	All other reportable incidents	47267	Single-chamber implantable pacemaker, rate-responsive	CT141	Electrodes/Leads and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery			
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery			
					CT145	Cardiovascular devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery			
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery			
					CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery			
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery			
					CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery			
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery			
					CT233	Surgical	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery			
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery			
					CT2406	Implantable	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery			
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery			
					CT2407	Active implantable	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery			

						F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery			
				CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery			
						F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery			
				CT2727	Active implant devices/ancillaries	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery			
						F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery			
				CT2750	Implantable cardiac electrical stimulators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery			
						F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery			
				CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery			
						F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery			
				CT2985	Active	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery			
						F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery			
				CT3001	Implantable metal	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery			
						F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery			
				CT3002	Implantable synthetic polymer	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery			
						F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery			
				CT3006	Implantable inorganic material	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery			
						F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery			
				CT314	Electrically powered	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery			
						F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery			
				CT321	Long-term surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery			
						F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery			
				CT449	Generators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery			
						F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery			
				CT981	Single use	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery			
						F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery			
				CT983	Surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery			
						F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery			
6900	2021	All other reportable incidents	35852	Single-chamber implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	14/10/2021	14/10/2021
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	14/10/2021	14/10/2021
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	14/10/2021	14/10/2021
6901	2021	All other reportable incidents	47265	Dual-chamber implantable pacemaker, rate-responsive	CT141	Electrodes/Leads and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT145	Cardiovascular devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT233	Surgical	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT2406	Implantable	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT2407	Active implantable	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT2727	Active implant devices/ancillaries	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT2750	Implantable cardiac electrical stimulators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT2985	Active	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT3001	Implantable metal	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT3002	Implantable synthetic polymer	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT3006	Implantable inorganic material	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT314	Electrically powered	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT321	Long-term surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT449	Generators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT981	Single use	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT983	Surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
6902	2021	All other reportable incidents	17882	Physiologic-monitoring defibrillation system	CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070803	Failure to Power Up		
							F27	Problem Identified During Non-Clinical Procedure	A070803	Failure to Power Up		
6903	2021	All other reportable incidents	17882	Physiologic-monitoring defibrillation system	CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070803	Failure to Power Up		
							F27	Problem Identified During Non-Clinical Procedure	A070803	Failure to Power Up		
6904	2022	All other reportable incidents	47265	Dual-chamber implantable pacemaker, rate-responsive	CT141	Electrodes/Leads and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	10/06/2022	10/06/2022
									A071102	Failure to Interrogate	10/06/2022	10/06/2022
							F1901	Additional Surgery	A070504	Premature Discharge of Battery	10/06/2022	10/06/2022
									A071102	Failure to Interrogate	10/06/2022	10/06/2022
					CT145	Cardiovascular devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	10/06/2022	10/06/2022
									A071102	Failure to Interrogate	10/06/2022	10/06/2022
							F1901	Additional Surgery	A070504	Premature Discharge of Battery	10/06/2022	10/06/2022
									A071102	Failure to Interrogate	10/06/2022	10/06/2022
					CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	10/06/2022	10/06/2022
									A071102	Failure to Interrogate	10/06/2022	10/06/2022
							F1901	Additional Surgery	A070504	Premature Discharge of Battery	10/06/2022	10/06/2022
									A071102	Failure to Interrogate	10/06/2022	10/06/2022
					CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	10/06/2022	10/06/2022
									A071102	Failure to Interrogate	10/06/2022	10/06/2022

							F1901	Additional Surgery	A070504	Premature Discharge of Battery	10/06/2022	10/06/2022
									A071102	Failure to Interrogate	10/06/2022	10/06/2022
					CT233	Surgical	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	10/06/2022	10/06/2022
									A071102	Failure to Interrogate	10/06/2022	10/06/2022
							F1901	Additional Surgery	A070504	Premature Discharge of Battery	10/06/2022	10/06/2022
									A071102	Failure to Interrogate	10/06/2022	10/06/2022
					CT2406	Implantable	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	10/06/2022	10/06/2022
									A071102	Failure to Interrogate	10/06/2022	10/06/2022
							F1901	Additional Surgery	A070504	Premature Discharge of Battery	10/06/2022	10/06/2022
									A071102	Failure to Interrogate	10/06/2022	10/06/2022
					CT2407	Active implantable	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	10/06/2022	10/06/2022
									A071102	Failure to Interrogate	10/06/2022	10/06/2022
							F1901	Additional Surgery	A070504	Premature Discharge of Battery	10/06/2022	10/06/2022
									A071102	Failure to Interrogate	10/06/2022	10/06/2022
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	10/06/2022	10/06/2022
									A071102	Failure to Interrogate	10/06/2022	10/06/2022
							F1901	Additional Surgery	A070504	Premature Discharge of Battery	10/06/2022	10/06/2022
									A071102	Failure to Interrogate	10/06/2022	10/06/2022
					CT2727	Active implant devices/ancillaries	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	10/06/2022	10/06/2022
									A071102	Failure to Interrogate	10/06/2022	10/06/2022
							F1901	Additional Surgery	A070504	Premature Discharge of Battery	10/06/2022	10/06/2022
									A071102	Failure to Interrogate	10/06/2022	10/06/2022
					CT2750	Implantable cardiac electrical stimulators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	10/06/2022	10/06/2022
									A071102	Failure to Interrogate	10/06/2022	10/06/2022
							F1901	Additional Surgery	A070504	Premature Discharge of Battery	10/06/2022	10/06/2022
									A071102	Failure to Interrogate	10/06/2022	10/06/2022
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	10/06/2022	10/06/2022
									A071102	Failure to Interrogate	10/06/2022	10/06/2022
							F1901	Additional Surgery	A070504	Premature Discharge of Battery	10/06/2022	10/06/2022
									A071102	Failure to Interrogate	10/06/2022	10/06/2022
					CT2985	Active	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	10/06/2022	10/06/2022
									A071102	Failure to Interrogate	10/06/2022	10/06/2022
							F1901	Additional Surgery	A070504	Premature Discharge of Battery	10/06/2022	10/06/2022
									A071102	Failure to Interrogate	10/06/2022	10/06/2022
					CT3001	Implantable metal	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	10/06/2022	10/06/2022
									A071102	Failure to Interrogate	10/06/2022	10/06/2022
							F1901	Additional Surgery	A070504	Premature Discharge of Battery	10/06/2022	10/06/2022
									A071102	Failure to Interrogate	10/06/2022	10/06/2022
					CT3002	Implantable synthetic polymer	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	10/06/2022	10/06/2022
									A071102	Failure to Interrogate	10/06/2022	10/06/2022
							F1901	Additional Surgery	A070504	Premature Discharge of Battery	10/06/2022	10/06/2022
									A071102	Failure to Interrogate	10/06/2022	10/06/2022
					CT3006	Implantable inorganic material	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	10/06/2022	10/06/2022
									A071102	Failure to Interrogate	10/06/2022	10/06/2022
							F1901	Additional Surgery	A070504	Premature Discharge of Battery	10/06/2022	10/06/2022
									A071102	Failure to Interrogate	10/06/2022	10/06/2022
					CT314	Electrically powered	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	10/06/2022	10/06/2022
									A071102	Failure to Interrogate	10/06/2022	10/06/2022
							F1901	Additional Surgery	A070504	Premature Discharge of Battery	10/06/2022	10/06/2022
									A071102	Failure to Interrogate	10/06/2022	10/06/2022
					CT321	Long-term surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	10/06/2022	10/06/2022
									A071102	Failure to Interrogate	10/06/2022	10/06/2022
							F1901	Additional Surgery	A070504	Premature Discharge of Battery	10/06/2022	10/06/2022
									A071102	Failure to Interrogate	10/06/2022	10/06/2022
					CT449	Generators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	10/06/2022	10/06/2022
									A071102	Failure to Interrogate	10/06/2022	10/06/2022
							F1901	Additional Surgery	A070504	Premature Discharge of Battery	10/06/2022	10/06/2022
									A071102	Failure to Interrogate	10/06/2022	10/06/2022
					CT981	Single use	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	10/06/2022	10/06/2022
									A071102	Failure to Interrogate	10/06/2022	10/06/2022
							F1901	Additional Surgery	A070504	Premature Discharge of Battery	10/06/2022	10/06/2022
									A071102	Failure to Interrogate	10/06/2022	10/06/2022
					CT983	Surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	10/06/2022	10/06/2022
									A071102	Failure to Interrogate	10/06/2022	10/06/2022
							F1901	Additional Surgery	A070504	Premature Discharge of Battery	10/06/2022	10/06/2022
									A071102	Failure to Interrogate	10/06/2022	10/06/2022
6905	2021	All other reportable incidents	17882	Physiologic-monitoring defibrillation system	CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070803	Failure to Power Up		
							F27	Problem Identified During Non-Clinical Procedure	A070803	Failure to Power Up		
6906	2022	All other reportable incidents	35223	Endocardial/interventricular septal pacing lead	CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A0701	Capturing Problem	03/06/2022	03/06/2022
									A0709	Device Sensing Problem	03/06/2022	03/06/2022
							F1901	Additional Surgery	A0701	Capturing Problem	03/06/2022	03/06/2022
									A0709	Device Sensing Problem	03/06/2022	03/06/2022
6907	2021	All other reportable incidents	17882	Physiologic-monitoring defibrillation system	CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070803	Failure to Power Up		
							F27	Problem Identified During Non-Clinical Procedure	A070803	Failure to Power Up		
6908	2021	All other reportable incidents	17882	Physiologic-monitoring defibrillation system	CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070803	Failure to Power Up		
							F27	Problem Identified During Non-Clinical Procedure	A070803	Failure to Power Up		
6909	2021	All other reportable incidents	17882	Physiologic-monitoring defibrillation system	CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070803	Failure to Power Up		
							F27	Problem Identified During Non-Clinical Procedure	A070803	Failure to Power Up		
6910	2022	All other reportable incidents	60789	Intracardiac pacemaker	CT130	Catheters and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT145	Cardiovascular devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT233	Surgical	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		

					CT2406	Implantable	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT2407	Active implantable	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT2727	Active implant devices/ancillaries	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT2750	Implantable cardiac electrical stimulators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT2818	Catheterization/Cannulation/Fluid path devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT2985	Active	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT3001	Implantable metal	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT3002	Implantable synthetic polymer	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT3006	Implantable inorganic material	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT314	Electrically powered	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT321	Long-term surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT981	Single use	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT983	Surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
6911	2021	All other reportable incidents	17882	Physiologic-monitoring defibrillation system	CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070803	Failure to Power Up		
							F27	Problem Identified During Non-Clinical Procedure	A070803	Failure to Power Up		
6912	2022	All other reportable incidents	47265	Dual-chamber implantable pacemaker, rate-responsive	CT141	Electrodes/Leads and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
					CT145	Cardiovascular devices	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
					CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
					CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
					CT233	Surgical	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
					CT2406	Implantable	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
					CT2407	Active implantable	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
					CT2727	Active implant devices/ancillaries	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
					CT2750	Implantable cardiac electrical stimulators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
					CT2985	Active	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
					CT3001	Implantable metal	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
					CT3002	Implantable synthetic polymer	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		

					CT3006	Implantable inorganic material	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem			
									A070504	Premature Discharge of Battery			
							F26	No Health Consequences or Impact	A0705	Battery Problem			
									A070504	Premature Discharge of Battery			
					CT314	Electrically powered	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem			
									A070504	Premature Discharge of Battery			
							F26	No Health Consequences or Impact	A0705	Battery Problem			
									A070504	Premature Discharge of Battery			
					CT321	Long-term surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem			
									A070504	Premature Discharge of Battery			
							F26	No Health Consequences or Impact	A0705	Battery Problem			
									A070504	Premature Discharge of Battery			
					CT449	Generators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem			
									A070504	Premature Discharge of Battery			
							F26	No Health Consequences or Impact	A0705	Battery Problem			
									A070504	Premature Discharge of Battery			
					CT981	Single use	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem			
									A070504	Premature Discharge of Battery			
							F26	No Health Consequences or Impact	A0705	Battery Problem			
									A070504	Premature Discharge of Battery			
					CT983	Surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem			
									A070504	Premature Discharge of Battery			
							F26	No Health Consequences or Impact	A0705	Battery Problem			
									A070504	Premature Discharge of Battery			
6913	2021	All other reportable incidents	17882	Physiologic-monitoring defibrillation system	CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070803	Failure to Power Up			
							F26	No Health Consequences or Impact	A070803	Failure to Power Up			
6914	2022	All other reportable incidents	35852	Single-chamber implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem	29/09/2022	29/09/2022	
									A070504	Premature Discharge of Battery	29/09/2022	29/09/2022	
							F08	Hospitalization or Prolonged Hospitalization	A0705	Battery Problem	29/09/2022	29/09/2022	
									A070504	Premature Discharge of Battery	29/09/2022	29/09/2022	
							F1905	Device Revision or Replacement	A0705	Battery Problem	29/09/2022	29/09/2022	
									A070504	Premature Discharge of Battery	29/09/2022	29/09/2022	
6915	2021	All other reportable incidents	35852	Single-chamber implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem	25/01/2022	25/01/2022	
									A070504	Premature Discharge of Battery	25/01/2022	25/01/2022	
							F08	Hospitalization or Prolonged Hospitalization	A0705	Battery Problem	25/01/2022	25/01/2022	
									A070504	Premature Discharge of Battery	25/01/2022	25/01/2022	
							F1903	Device Explantation	A0705	Battery Problem	25/01/2022	25/01/2022	
									A070504	Premature Discharge of Battery	25/01/2022	25/01/2022	
6916	2021	All other reportable incidents	36241	Pacemaker/defibrillator lead	CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A072201	High Impedance			
							F26	No Health Consequences or Impact	A072201	High Impedance			
6917	2022	All other reportable incidents	47263	Cardiac resynchronization therapy implantable pacemaker	CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	30/05/2022	30/05/2022	
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	30/05/2022	30/05/2022	
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	30/05/2022	30/05/2022	
					CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	30/05/2022	30/05/2022	
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	30/05/2022	30/05/2022	
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	30/05/2022	30/05/2022	
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	30/05/2022	30/05/2022	
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	30/05/2022	30/05/2022	
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	30/05/2022	30/05/2022	
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	30/05/2022	30/05/2022	
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	30/05/2022	30/05/2022	
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	30/05/2022	30/05/2022	
6918	2022	All other reportable incidents	17882	Physiologic-monitoring defibrillation system	CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A040101	Fracture			
							F27	Problem Identified During Non-Clinical Procedure	A040101	Fracture			
6919	2021	All other reportable incidents	35223	Endocardial/interventricular septal pacing lead	CT1684	Cardiac pacemakers and associated devices	E0112	Dizziness	A0401	Break			
									A070101	Failure to Capture			
									A070909	Oversensing			
									A072202	Low Impedance			
									A090407	Therapy Delivered to Incorrect Body Area			
							E2403	No Clinical Signs, Symptoms or Conditions	A0401	Break			
									A070101	Failure to Capture			
									A070909	Oversensing			
									A072202	Low Impedance			
									A090407	Therapy Delivered to Incorrect Body Area			
							F1901	Additional Surgery	A0401	Break			
									A070101	Failure to Capture			
									A070909	Oversensing			
									A072202	Low Impedance			
									A090407	Therapy Delivered to Incorrect Body Area			
6920	2022	All other reportable incidents	35223	Endocardial/interventricular septal pacing lead	CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A0401	Break			
									A0709	Device Sensing Problem			
									A070909	Oversensing			
							F1901	Additional Surgery	A0401	Break			
									A0709	Device Sensing Problem			
									A070909	Oversensing			
6921	2022	All other reportable incidents	35852	Single-chamber implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E2104	Electric Shock	A070909	Oversensing			
									A070910	Undersensing			
									A071302	Inappropriate/Inadequate Shock/Stimulation			
									A1301	Failure to Read Input Signal			
							E2403	No Clinical Signs, Symptoms or Conditions	A070909	Oversensing			
									A070910	Undersensing			
									A071302	Inappropriate/Inadequate Shock/Stimulation			
									A1301	Failure to Read Input Signal			
							F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A070909	Oversensing			
									A070910	Undersensing			
									A071302	Inappropriate/Inadequate Shock/Stimulation			
									A1301	Failure to Read Input Signal			
6922	2021	All other reportable incidents	17882	Physiologic-monitoring defibrillation system	CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070803	Failure to Power Up			

							F26	No Health Consequences or Impact	A070803	Failure to Power Up		
6923	2021	All other reportable incidents	37265	Dual-chamber implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A070909	Oversensing		
							F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A070909	Oversensing		
6924	2021	All other reportable incidents	17882	Physiologic-monitoring defibrillation system	CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A07	Electrical /Electronic Property Problem		
							F27	Problem Identified During Non-Clinical Procedure	A07	Electrical /Electronic Property Problem		
6925	2021	All other reportable incidents	47263	Cardiac resynchronization therapy implantable pacemaker	CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A0908	Incorrect, Inadequate or Imprecise Result or Readings	21/10/2021	21/10/2021
							F08	Hospitalization or Prolonged Hospitalization	A0908	Incorrect, Inadequate or Imprecise Result or Readings	21/10/2021	21/10/2021
							F1903	Device Explantation	A0908	Incorrect, Inadequate or Imprecise Result or Readings	21/10/2021	21/10/2021
					CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A0908	Incorrect, Inadequate or Imprecise Result or Readings	21/10/2021	21/10/2021
							F08	Hospitalization or Prolonged Hospitalization	A0908	Incorrect, Inadequate or Imprecise Result or Readings	21/10/2021	21/10/2021
							F1903	Device Explantation	A0908	Incorrect, Inadequate or Imprecise Result or Readings	21/10/2021	21/10/2021
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A0908	Incorrect, Inadequate or Imprecise Result or Readings	21/10/2021	21/10/2021
							F08	Hospitalization or Prolonged Hospitalization	A0908	Incorrect, Inadequate or Imprecise Result or Readings	21/10/2021	21/10/2021
							F1903	Device Explantation	A0908	Incorrect, Inadequate or Imprecise Result or Readings	21/10/2021	21/10/2021
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A0908	Incorrect, Inadequate or Imprecise Result or Readings	21/10/2021	21/10/2021
							F08	Hospitalization or Prolonged Hospitalization	A0908	Incorrect, Inadequate or Imprecise Result or Readings	21/10/2021	21/10/2021
							F1903	Device Explantation	A0908	Incorrect, Inadequate or Imprecise Result or Readings	21/10/2021	21/10/2021
6926	2021	All other reportable incidents	47265	Dual-chamber implantable pacemaker, rate-responsive	CT141	Electrodes/Leads and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	20/10/2021	20/10/2021
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	20/10/2021	20/10/2021
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	20/10/2021	20/10/2021
					CT145	Cardiovascular devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	20/10/2021	20/10/2021
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	20/10/2021	20/10/2021
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	20/10/2021	20/10/2021
					CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	20/10/2021	20/10/2021
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	20/10/2021	20/10/2021
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	20/10/2021	20/10/2021
					CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	20/10/2021	20/10/2021
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	20/10/2021	20/10/2021
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	20/10/2021	20/10/2021
					CT233	Surgical	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	20/10/2021	20/10/2021
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	20/10/2021	20/10/2021
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	20/10/2021	20/10/2021
					CT2406	Implantable	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	20/10/2021	20/10/2021
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	20/10/2021	20/10/2021
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	20/10/2021	20/10/2021
					CT2407	Active implantable	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	20/10/2021	20/10/2021
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	20/10/2021	20/10/2021
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	20/10/2021	20/10/2021
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	20/10/2021	20/10/2021
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	20/10/2021	20/10/2021
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	20/10/2021	20/10/2021
					CT2727	Active implant devices/ancillaries	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	20/10/2021	20/10/2021
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	20/10/2021	20/10/2021
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	20/10/2021	20/10/2021
					CT2750	Implantable cardiac electrical stimulators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	20/10/2021	20/10/2021
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	20/10/2021	20/10/2021
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	20/10/2021	20/10/2021
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	20/10/2021	20/10/2021
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	20/10/2021	20/10/2021
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	20/10/2021	20/10/2021
					CT2985	Active	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	20/10/2021	20/10/2021
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	20/10/2021	20/10/2021
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	20/10/2021	20/10/2021
					CT3001	Implantable metal	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	20/10/2021	20/10/2021
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	20/10/2021	20/10/2021
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	20/10/2021	20/10/2021
					CT3002	Implantable synthetic polymer	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	20/10/2021	20/10/2021
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	20/10/2021	20/10/2021
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	20/10/2021	20/10/2021
					CT3006	Implantable inorganic material	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	20/10/2021	20/10/2021
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	20/10/2021	20/10/2021
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	20/10/2021	20/10/2021
					CT314	Electrically powered	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	20/10/2021	20/10/2021
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	20/10/2021	20/10/2021
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	20/10/2021	20/10/2021
					CT321	Long-term surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	20/10/2021	20/10/2021
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	20/10/2021	20/10/2021
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	20/10/2021	20/10/2021
					CT449	Generators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	20/10/2021	20/10/2021
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	20/10/2021	20/10/2021
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	20/10/2021	20/10/2021
					CT981	Single use	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	20/10/2021	20/10/2021
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	20/10/2021	20/10/2021
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	20/10/2021	20/10/2021
					CT983	Surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	20/10/2021	20/10/2021
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	20/10/2021	20/10/2021
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	20/10/2021	20/10/2021
6927	2021	All other reportable incidents	37265	Dual-chamber implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E2104	Electric Shock	A0712	Pacing Problem		
							E2403	No Clinical Signs, Symptoms or Conditions	A0712	Pacing Problem		

							F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A0712	Pacing Problem		
6928	2022	All other reportable incidents	47265	Dual-chamber implantable pacemaker, rate-responsive	CT141	Electrodes/Leads and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT145	Cardiovascular devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT233	Surgical	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT2406	Implantable	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT2407	Active implantable	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT2727	Active implant devices/ancillaries	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT2750	Implantable cardiac electrical stimulators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT2985	Active	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT3001	Implantable metal	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT3002	Implantable synthetic polymer	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT3006	Implantable inorganic material	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT314	Electrically powered	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT321	Long-term surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT449	Generators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT981	Single use	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT983	Surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
6929	2021	All other reportable incidents	35223	Endocardial/interventricular septal pacing lead	CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A0401	Break		
									A070909	Oversensing		
							F1901	Additional Surgery	A0401	Break		
									A070909	Oversensing		
6930	2021	All other reportable incidents	17882	Physiologic-monitoring defibrillation system	CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070912	Incorrect Interpretation of Signal		
							F27	Problem Identified During Non-Clinical Procedure	A070912	Incorrect Interpretation of Signal		
6931	2021	All other reportable incidents	47265	Dual-chamber implantable pacemaker, rate-responsive	CT141	Electrodes/Leads and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	03/11/2021	03/11/2021
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	03/11/2021	03/11/2021
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	03/11/2021	03/11/2021
					CT145	Cardiovascular devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	03/11/2021	03/11/2021
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	03/11/2021	03/11/2021
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	03/11/2021	03/11/2021
					CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	03/11/2021	03/11/2021
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	03/11/2021	03/11/2021
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	03/11/2021	03/11/2021
					CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	03/11/2021	03/11/2021
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	03/11/2021	03/11/2021
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	03/11/2021	03/11/2021
					CT233	Surgical	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	03/11/2021	03/11/2021
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	03/11/2021	03/11/2021
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	03/11/2021	03/11/2021
					CT2406	Implantable	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	03/11/2021	03/11/2021
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	03/11/2021	03/11/2021
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	03/11/2021	03/11/2021
					CT2407	Active implantable	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	03/11/2021	03/11/2021
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	03/11/2021	03/11/2021
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	03/11/2021	03/11/2021
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	03/11/2021	03/11/2021
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	03/11/2021	03/11/2021
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	03/11/2021	03/11/2021
					CT2727	Active implant devices/ancillaries	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	03/11/2021	03/11/2021
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	03/11/2021	03/11/2021
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	03/11/2021	03/11/2021
					CT2750	Implantable cardiac electrical stimulators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	03/11/2021	03/11/2021
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	03/11/2021	03/11/2021
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	03/11/2021	03/11/2021
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	03/11/2021	03/11/2021
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	03/11/2021	03/11/2021
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	03/11/2021	03/11/2021
					CT2985	Active	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	03/11/2021	03/11/2021
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	03/11/2021	03/11/2021
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	03/11/2021	03/11/2021
					CT3001	Implantable metal	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	03/11/2021	03/11/2021
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	03/11/2021	03/11/2021
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	03/11/2021	03/11/2021
					CT3002	Implantable synthetic polymer	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	03/11/2021	03/11/2021
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	03/11/2021	03/11/2021

						F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	03/11/2021	03/11/2021	
				CT3006	Implantable inorganic material	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	03/11/2021	03/11/2021	
						F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	03/11/2021	03/11/2021	
						F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	03/11/2021	03/11/2021	
				CT314	Electrically powered	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	03/11/2021	03/11/2021	
						F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	03/11/2021	03/11/2021	
						F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	03/11/2021	03/11/2021	
				CT321	Long-term surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	03/11/2021	03/11/2021	
						F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	03/11/2021	03/11/2021	
						F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	03/11/2021	03/11/2021	
				CT449	Generators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	03/11/2021	03/11/2021	
						F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	03/11/2021	03/11/2021	
						F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	03/11/2021	03/11/2021	
				CT981	Single use	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	03/11/2021	03/11/2021	
						F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	03/11/2021	03/11/2021	
						F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	03/11/2021	03/11/2021	
				CT983	Surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	03/11/2021	03/11/2021	
						F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	03/11/2021	03/11/2021	
						F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	03/11/2021	03/11/2021	
6932	2021	All other reportable incidents	35852	Single-chamber implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem	04/11/2021	04/11/2021
									A070504	Premature Discharge of Battery	04/11/2021	04/11/2021
							F08	Hospitalization or Prolonged Hospitalization	A0705	Battery Problem	04/11/2021	04/11/2021
									A070504	Premature Discharge of Battery	04/11/2021	04/11/2021
							F1905	Device Revision or Replacement	A0705	Battery Problem	04/11/2021	04/11/2021
									A070504	Premature Discharge of Battery	04/11/2021	04/11/2021
6933	2021	All other reportable incidents	47265	Dual-chamber implantable pacemaker, rate-responsive	CT141	Electrodes/Leads and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	18/10/2021	18/10/2021
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	18/10/2021	18/10/2021
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	18/10/2021	18/10/2021
					CT145	Cardiovascular devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	18/10/2021	18/10/2021
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	18/10/2021	18/10/2021
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	18/10/2021	18/10/2021
					CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	18/10/2021	18/10/2021
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	18/10/2021	18/10/2021
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	18/10/2021	18/10/2021
					CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	18/10/2021	18/10/2021
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	18/10/2021	18/10/2021
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	18/10/2021	18/10/2021
					CT233	Surgical	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	18/10/2021	18/10/2021
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	18/10/2021	18/10/2021
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	18/10/2021	18/10/2021
					CT2406	Implantable	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	18/10/2021	18/10/2021
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	18/10/2021	18/10/2021
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	18/10/2021	18/10/2021
					CT2407	Active implantable	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	18/10/2021	18/10/2021
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	18/10/2021	18/10/2021
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	18/10/2021	18/10/2021
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	18/10/2021	18/10/2021
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	18/10/2021	18/10/2021
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	18/10/2021	18/10/2021
					CT2727	Active implant devices/ancillaries	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	18/10/2021	18/10/2021
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	18/10/2021	18/10/2021
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	18/10/2021	18/10/2021
					CT2750	Implantable cardiac electrical stimulators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	18/10/2021	18/10/2021
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	18/10/2021	18/10/2021
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	18/10/2021	18/10/2021
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	18/10/2021	18/10/2021
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	18/10/2021	18/10/2021
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	18/10/2021	18/10/2021
					CT2985	Active	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	18/10/2021	18/10/2021
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	18/10/2021	18/10/2021
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	18/10/2021	18/10/2021
					CT3001	Implantable metal	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	18/10/2021	18/10/2021
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	18/10/2021	18/10/2021
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	18/10/2021	18/10/2021
					CT3002	Implantable synthetic polymer	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	18/10/2021	18/10/2021
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	18/10/2021	18/10/2021
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	18/10/2021	18/10/2021
					CT3006	Implantable inorganic material	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	18/10/2021	18/10/2021
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	18/10/2021	18/10/2021
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	18/10/2021	18/10/2021
					CT314	Electrically powered	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	18/10/2021	18/10/2021
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	18/10/2021	18/10/2021
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	18/10/2021	18/10/2021
					CT321	Long-term surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	18/10/2021	18/10/2021
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	18/10/2021	18/10/2021
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	18/10/2021	18/10/2021
					CT449	Generators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	18/10/2021	18/10/2021
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	18/10/2021	18/10/2021
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	18/10/2021	18/10/2021
					CT981	Single use	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	18/10/2021	18/10/2021
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	18/10/2021	18/10/2021
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	18/10/2021	18/10/2021
					CT983	Surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	18/10/2021	18/10/2021
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	18/10/2021	18/10/2021
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	18/10/2021	18/10/2021
6934	2021	All other reportable incidents	35852	Single-chamber implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E2104	Electric Shock	A071302	Inappropriate/Inadequate Shock/Stimulation		
							E2403	No Clinical Signs, Symptoms or Conditions	A071302	Inappropriate/Inadequate Shock/Stimulation		
							F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A071302	Inappropriate/Inadequate Shock/Stimulation		

6935	2021	All other reportable incidents	35852	Single-chamber implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem	16/12/2021	16/12/2021
									A070504	Premature Discharge of Battery	16/12/2021	16/12/2021
							F08	Hospitalization or Prolonged Hospitalization	A0705	Battery Problem	16/12/2021	16/12/2021
									A070504	Premature Discharge of Battery	16/12/2021	16/12/2021
							F1905	Device Revision or Replacement	A0705	Battery Problem	16/12/2021	16/12/2021
									A070504	Premature Discharge of Battery	16/12/2021	16/12/2021
6936	2021	All other reportable incidents	60910	Coronary venous pacing lead	CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070102	High Capture Threshold		
									A072202	Low Impedance		
							F26	No Health Consequences or Impact	A070102	High Capture Threshold		
									A072202	Low Impedance		
6937	2021	All other reportable incidents	47265	Dual-chamber implantable pacemaker, rate-responsive	CT141	Electrodes/Leads and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	27/03/2022	27/03/2022
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	27/03/2022	27/03/2022
							F1903	Device Explantation	A070504	Premature Discharge of Battery	27/03/2022	27/03/2022
					CT145	Cardiovascular devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	27/03/2022	27/03/2022
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	27/03/2022	27/03/2022
							F1903	Device Explantation	A070504	Premature Discharge of Battery	27/03/2022	27/03/2022
					CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	27/03/2022	27/03/2022
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	27/03/2022	27/03/2022
							F1903	Device Explantation	A070504	Premature Discharge of Battery	27/03/2022	27/03/2022
					CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	27/03/2022	27/03/2022
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	27/03/2022	27/03/2022
							F1903	Device Explantation	A070504	Premature Discharge of Battery	27/03/2022	27/03/2022
					CT233	Surgical	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	27/03/2022	27/03/2022
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	27/03/2022	27/03/2022
							F1903	Device Explantation	A070504	Premature Discharge of Battery	27/03/2022	27/03/2022
					CT2406	Implantable	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	27/03/2022	27/03/2022
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	27/03/2022	27/03/2022
							F1903	Device Explantation	A070504	Premature Discharge of Battery	27/03/2022	27/03/2022
					CT2407	Active implantable	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	27/03/2022	27/03/2022
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	27/03/2022	27/03/2022
							F1903	Device Explantation	A070504	Premature Discharge of Battery	27/03/2022	27/03/2022
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	27/03/2022	27/03/2022
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	27/03/2022	27/03/2022
							F1903	Device Explantation	A070504	Premature Discharge of Battery	27/03/2022	27/03/2022
					CT2727	Active implant devices/ancillaries	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	27/03/2022	27/03/2022
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	27/03/2022	27/03/2022
							F1903	Device Explantation	A070504	Premature Discharge of Battery	27/03/2022	27/03/2022
					CT2750	Implantable cardiac electrical stimulators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	27/03/2022	27/03/2022
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	27/03/2022	27/03/2022
							F1903	Device Explantation	A070504	Premature Discharge of Battery	27/03/2022	27/03/2022
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	27/03/2022	27/03/2022
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	27/03/2022	27/03/2022
							F1903	Device Explantation	A070504	Premature Discharge of Battery	27/03/2022	27/03/2022
					CT2985	Active	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	27/03/2022	27/03/2022
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	27/03/2022	27/03/2022
							F1903	Device Explantation	A070504	Premature Discharge of Battery	27/03/2022	27/03/2022
					CT3001	Implantable metal	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	27/03/2022	27/03/2022
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	27/03/2022	27/03/2022
							F1903	Device Explantation	A070504	Premature Discharge of Battery	27/03/2022	27/03/2022
					CT3002	Implantable synthetic polymer	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	27/03/2022	27/03/2022
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	27/03/2022	27/03/2022
							F1903	Device Explantation	A070504	Premature Discharge of Battery	27/03/2022	27/03/2022
					CT3006	Implantable inorganic material	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	27/03/2022	27/03/2022
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	27/03/2022	27/03/2022
							F1903	Device Explantation	A070504	Premature Discharge of Battery	27/03/2022	27/03/2022
					CT314	Electrically powered	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	27/03/2022	27/03/2022
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	27/03/2022	27/03/2022
							F1903	Device Explantation	A070504	Premature Discharge of Battery	27/03/2022	27/03/2022
					CT321	Long-term surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	27/03/2022	27/03/2022
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	27/03/2022	27/03/2022
							F1903	Device Explantation	A070504	Premature Discharge of Battery	27/03/2022	27/03/2022
					CT449	Generators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	27/03/2022	27/03/2022
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	27/03/2022	27/03/2022
							F1903	Device Explantation	A070504	Premature Discharge of Battery	27/03/2022	27/03/2022
					CT981	Single use	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	27/03/2022	27/03/2022
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	27/03/2022	27/03/2022
							F1903	Device Explantation	A070504	Premature Discharge of Battery	27/03/2022	27/03/2022
					CT983	Surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	27/03/2022	27/03/2022
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	27/03/2022	27/03/2022
							F1903	Device Explantation	A070504	Premature Discharge of Battery	27/03/2022	27/03/2022
6938	2022	All other reportable incidents	35852	Single-chamber implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem	12/07/2022	12/07/2022
									A070504	Premature Discharge of Battery	12/07/2022	12/07/2022
							F08	Hospitalization or Prolonged Hospitalization	A0705	Battery Problem	12/07/2022	12/07/2022
									A070504	Premature Discharge of Battery	12/07/2022	12/07/2022
							F1905	Device Revision or Replacement	A0705	Battery Problem	12/07/2022	12/07/2022
									A070504	Premature Discharge of Battery	12/07/2022	12/07/2022
6939	2022	All other reportable incidents	36241	Pacemaker/defibrillator lead	CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A072201	High Impedance		
							F26	No Health Consequences or Impact	A072201	High Impedance		
6940	2021	All other reportable incidents	35852	Single-chamber implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem	09/11/2021	09/11/2021
									A070504	Premature Discharge of Battery	09/11/2021	09/11/2021
							F08	Hospitalization or Prolonged Hospitalization	A0705	Battery Problem	09/11/2021	09/11/2021
									A070504	Premature Discharge of Battery	09/11/2021	09/11/2021
							F1905	Device Revision or Replacement	A0705	Battery Problem	09/11/2021	09/11/2021
									A070504	Premature Discharge of Battery	09/11/2021	09/11/2021
6941	2021	All other reportable incidents	47265	Dual-chamber implantable pacemaker, rate-responsive	CT141	Electrodes/Leads and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070101	Failure to Capture	21/10/2021	21/10/2021
							F1901	Additional Surgery	A070101	Failure to Capture	21/10/2021	21/10/2021
					CT145	Cardiovascular devices	E2403	No Clinical Signs, Symptoms or Conditions	A070101	Failure to Capture	21/10/2021	21/10/2021
							F1901	Additional Surgery	A070101	Failure to Capture	21/10/2021	21/10/2021

					CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070101	Failure to Capture	21/10/2021	21/10/2021
							F1901	Additional Surgery	A070101	Failure to Capture	21/10/2021	21/10/2021
					CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A070101	Failure to Capture	21/10/2021	21/10/2021
							F1901	Additional Surgery	A070101	Failure to Capture	21/10/2021	21/10/2021
					CT233	Surgical	E2403	No Clinical Signs, Symptoms or Conditions	A070101	Failure to Capture	21/10/2021	21/10/2021
							F1901	Additional Surgery	A070101	Failure to Capture	21/10/2021	21/10/2021
					CT2406	Implantable	E2403	No Clinical Signs, Symptoms or Conditions	A070101	Failure to Capture	21/10/2021	21/10/2021
							F1901	Additional Surgery	A070101	Failure to Capture	21/10/2021	21/10/2021
					CT2407	Active implantable	E2403	No Clinical Signs, Symptoms or Conditions	A070101	Failure to Capture	21/10/2021	21/10/2021
							F1901	Additional Surgery	A070101	Failure to Capture	21/10/2021	21/10/2021
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A070101	Failure to Capture	21/10/2021	21/10/2021
							F1901	Additional Surgery	A070101	Failure to Capture	21/10/2021	21/10/2021
					CT2727	Active implant devices/ancillaries	E2403	No Clinical Signs, Symptoms or Conditions	A070101	Failure to Capture	21/10/2021	21/10/2021
							F1901	Additional Surgery	A070101	Failure to Capture	21/10/2021	21/10/2021
					CT2750	Implantable cardiac electrical stimulators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070101	Failure to Capture	21/10/2021	21/10/2021
							F1901	Additional Surgery	A070101	Failure to Capture	21/10/2021	21/10/2021
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A070101	Failure to Capture	21/10/2021	21/10/2021
							F1901	Additional Surgery	A070101	Failure to Capture	21/10/2021	21/10/2021
					CT2985	Active	E2403	No Clinical Signs, Symptoms or Conditions	A070101	Failure to Capture	21/10/2021	21/10/2021
							F1901	Additional Surgery	A070101	Failure to Capture	21/10/2021	21/10/2021
					CT3001	Implantable metal	E2403	No Clinical Signs, Symptoms or Conditions	A070101	Failure to Capture	21/10/2021	21/10/2021
							F1901	Additional Surgery	A070101	Failure to Capture	21/10/2021	21/10/2021
					CT3002	Implantable synthetic polymer	E2403	No Clinical Signs, Symptoms or Conditions	A070101	Failure to Capture	21/10/2021	21/10/2021
							F1901	Additional Surgery	A070101	Failure to Capture	21/10/2021	21/10/2021
					CT3006	Implantable inorganic material	E2403	No Clinical Signs, Symptoms or Conditions	A070101	Failure to Capture	21/10/2021	21/10/2021
							F1901	Additional Surgery	A070101	Failure to Capture	21/10/2021	21/10/2021
					CT314	Electrically powered	E2403	No Clinical Signs, Symptoms or Conditions	A070101	Failure to Capture	21/10/2021	21/10/2021
							F1901	Additional Surgery	A070101	Failure to Capture	21/10/2021	21/10/2021
					CT321	Long-term surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A070101	Failure to Capture	21/10/2021	21/10/2021
							F1901	Additional Surgery	A070101	Failure to Capture	21/10/2021	21/10/2021
					CT449	Generators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070101	Failure to Capture	21/10/2021	21/10/2021
							F1901	Additional Surgery	A070101	Failure to Capture	21/10/2021	21/10/2021
					CT981	Single use	E2403	No Clinical Signs, Symptoms or Conditions	A070101	Failure to Capture	21/10/2021	21/10/2021
							F1901	Additional Surgery	A070101	Failure to Capture	21/10/2021	21/10/2021
					CT983	Surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A070101	Failure to Capture	21/10/2021	21/10/2021
							F1901	Additional Surgery	A070101	Failure to Capture	21/10/2021	21/10/2021
6942	2022	All other reportable incidents	47265	Dual-chamber implantable pacemaker, rate-responsive	CT141	Electrodes/Leads and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	24/08/2022	24/08/2022
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	24/08/2022	24/08/2022
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	24/08/2022	24/08/2022
					CT145	Cardiovascular devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	24/08/2022	24/08/2022
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	24/08/2022	24/08/2022
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	24/08/2022	24/08/2022
					CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	24/08/2022	24/08/2022
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	24/08/2022	24/08/2022
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	24/08/2022	24/08/2022
					CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	24/08/2022	24/08/2022
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	24/08/2022	24/08/2022
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	24/08/2022	24/08/2022
					CT233	Surgical	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	24/08/2022	24/08/2022
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	24/08/2022	24/08/2022
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	24/08/2022	24/08/2022
					CT2406	Implantable	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	24/08/2022	24/08/2022
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	24/08/2022	24/08/2022
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	24/08/2022	24/08/2022
					CT2407	Active implantable	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	24/08/2022	24/08/2022
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	24/08/2022	24/08/2022
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	24/08/2022	24/08/2022
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	24/08/2022	24/08/2022
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	24/08/2022	24/08/2022
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	24/08/2022	24/08/2022
					CT2727	Active implant devices/ancillaries	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	24/08/2022	24/08/2022
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	24/08/2022	24/08/2022
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	24/08/2022	24/08/2022
					CT2750	Implantable cardiac electrical stimulators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	24/08/2022	24/08/2022
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	24/08/2022	24/08/2022
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	24/08/2022	24/08/2022
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	24/08/2022	24/08/2022
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	24/08/2022	24/08/2022
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	24/08/2022	24/08/2022
					CT2985	Active	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	24/08/2022	24/08/2022
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	24/08/2022	24/08/2022
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	24/08/2022	24/08/2022
					CT3001	Implantable metal	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	24/08/2022	24/08/2022
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	24/08/2022	24/08/2022
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	24/08/2022	24/08/2022
					CT3002	Implantable synthetic polymer	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	24/08/2022	24/08/2022
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	24/08/2022	24/08/2022
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	24/08/2022	24/08/2022
					CT3006	Implantable inorganic material	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	24/08/2022	24/08/2022
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	24/08/2022	24/08/2022
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	24/08/2022	24/08/2022
					CT314	Electrically powered	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	24/08/2022	24/08/2022
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	24/08/2022	24/08/2022
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	24/08/2022	24/08/2022
					CT321	Long-term surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	24/08/2022	24/08/2022
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	24/08/2022	24/08/2022
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	24/08/2022	24/08/2022
					CT449	Generators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	24/08/2022	24/08/2022

							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	24/08/2022	24/08/2022
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	24/08/2022	24/08/2022
					CT981	Single use	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	24/08/2022	24/08/2022
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	24/08/2022	24/08/2022
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	24/08/2022	24/08/2022
					CT983	Surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	24/08/2022	24/08/2022
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	24/08/2022	24/08/2022
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	24/08/2022	24/08/2022
6943	2022	All other reportable incidents	36241	Pacemaker/defibrillator lead	CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070909	Oversensing		
							F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A070909	Oversensing		
6944	2022	All other reportable incidents	36241	Pacemaker/defibrillator lead	CT1684	Cardiac pacemakers and associated devices	E2104	Electric Shock	A071302	Inappropriate/Inadequate Shock/Stimulation		
									A072201	High Impedance		
							E2403	No Clinical Signs, Symptoms or Conditions	A071302	Inappropriate/Inadequate Shock/Stimulation		
									A072201	High Impedance		
							F08	Hospitalization or Prolonged Hospitalization	A071302	Inappropriate/Inadequate Shock/Stimulation		
									A072201	High Impedance		
							F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A071302	Inappropriate/Inadequate Shock/Stimulation		
									A072201	High Impedance		
							F2203	Imaging Required	A071302	Inappropriate/Inadequate Shock/Stimulation		
									A072201	High Impedance		
							F23	Unexpected Medical Intervention	A071302	Inappropriate/Inadequate Shock/Stimulation		
									A072201	High Impedance		
6945	2022	All other reportable incidents	37265	Dual-chamber implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
6946	2021	All other reportable incidents	17882	Physiologic-monitoring defibrillation system	CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070802	Intermittent Loss of Power		
							F26	No Health Consequences or Impact	A070802	Intermittent Loss of Power		
6947	2021	All other reportable incidents	17882	Physiologic-monitoring defibrillation system	CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070603	Failure to Charge		
							F27	Problem Identified During Non-Clinical Procedure	A070603	Failure to Charge		
6948	2021	All other reportable incidents	17882	Physiologic-monitoring defibrillation system	CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A0721	Circuit Failure		
							F27	Problem Identified During Non-Clinical Procedure	A0721	Circuit Failure		
6949	2021	All other reportable incidents	47270	Cardiac resynchronization therapy implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A0712	Pacing Problem		
							F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A0712	Pacing Problem		
6950	2021	All other reportable incidents	35852	Single-chamber implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem	19/11/2021	19/11/2021
									A070504	Premature Discharge of Battery	19/11/2021	19/11/2021
							F08	Hospitalization or Prolonged Hospitalization	A0705	Battery Problem	19/11/2021	19/11/2021
									A070504	Premature Discharge of Battery	19/11/2021	19/11/2021
							F1903	Device Explantation	A0705	Battery Problem	19/11/2021	19/11/2021
									A070504	Premature Discharge of Battery	19/11/2021	19/11/2021
6951	2021	All other reportable incidents	36241	Pacemaker/defibrillator lead	CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A072201	High Impedance		
							F26	No Health Consequences or Impact	A072201	High Impedance		
6952	2021	All other reportable incidents	35852	Single-chamber implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem	10/11/2021	10/11/2021
									A070504	Premature Discharge of Battery	10/11/2021	10/11/2021
							F08	Hospitalization or Prolonged Hospitalization	A0705	Battery Problem	10/11/2021	10/11/2021
									A070504	Premature Discharge of Battery	10/11/2021	10/11/2021
							F1905	Device Revision or Replacement	A0705	Battery Problem	10/11/2021	10/11/2021
									A070504	Premature Discharge of Battery	10/11/2021	10/11/2021
6953	2021	All other reportable incidents	35852	Single-chamber implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem	11/11/2021	11/11/2021
									A070504	Premature Discharge of Battery	11/11/2021	11/11/2021
							F08	Hospitalization or Prolonged Hospitalization	A0705	Battery Problem	11/11/2021	11/11/2021
									A070504	Premature Discharge of Battery	11/11/2021	11/11/2021
							F1905	Device Revision or Replacement	A0705	Battery Problem	11/11/2021	11/11/2021
									A070504	Premature Discharge of Battery	11/11/2021	11/11/2021
6954	2022	All other reportable incidents	17882	Physiologic-monitoring defibrillation system	CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A071302	Inappropriate/Inadequate Shock/Stimulation		
							F27	Problem Identified During Non-Clinical Procedure	A071302	Inappropriate/Inadequate Shock/Stimulation		
6955	2022	All other reportable incidents	35223	Endocardial/interventricular septal pacing lead	CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070910	Undersensing		
									A0722	Impedance Problem		
									A072202	Low Impedance		
							F23	Unexpected Medical Intervention	A070910	Undersensing		
									A0722	Impedance Problem		
									A072202	Low Impedance		
							F26	No Health Consequences or Impact	A070910	Undersensing		
									A0722	Impedance Problem		
									A072202	Low Impedance		
6956	2022	All other reportable incidents	47263	Cardiac resynchronization therapy implantable pacemaker	CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A0908	Incorrect, Inadequate or Imprecise Result or Readings	08/06/2022	08/06/2022
							F08	Hospitalization or Prolonged Hospitalization	A0908	Incorrect, Inadequate or Imprecise Result or Readings	08/06/2022	08/06/2022
							F1905	Device Revision or Replacement	A0908	Incorrect, Inadequate or Imprecise Result or Readings	08/06/2022	08/06/2022
					CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A0908	Incorrect, Inadequate or Imprecise Result or Readings	08/06/2022	08/06/2022
							F08	Hospitalization or Prolonged Hospitalization	A0908	Incorrect, Inadequate or Imprecise Result or Readings	08/06/2022	08/06/2022
							F1905	Device Revision or Replacement	A0908	Incorrect, Inadequate or Imprecise Result or Readings	08/06/2022	08/06/2022
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A0908	Incorrect, Inadequate or Imprecise Result or Readings	08/06/2022	08/06/2022
							F08	Hospitalization or Prolonged Hospitalization	A0908	Incorrect, Inadequate or Imprecise Result or Readings	08/06/2022	08/06/2022
							F1905	Device Revision or Replacement	A0908	Incorrect, Inadequate or Imprecise Result or Readings	08/06/2022	08/06/2022
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A0908	Incorrect, Inadequate or Imprecise Result or Readings	08/06/2022	08/06/2022
							F08	Hospitalization or Prolonged Hospitalization	A0908	Incorrect, Inadequate or Imprecise Result or Readings	08/06/2022	08/06/2022
							F1905	Device Revision or Replacement	A0908	Incorrect, Inadequate or Imprecise Result or Readings	08/06/2022	08/06/2022
6957	2021	All other reportable incidents	35223	Endocardial/interventricular septal pacing lead	CT1684	Cardiac pacemakers and associated devices	E0112	Dizziness	A040101	Fracture		
									A070103	Intermittent Capture		
							E2403	No Clinical Signs, Symptoms or Conditions	A040101	Fracture		
									A070103	Intermittent Capture		

							F23	Unexpected Medical Intervention	A040101	Fracture		
									A070103	Intermittent Capture		
6958	2021	All other reportable incidents	35852	Single-chamber implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A0203	Defective Device	25/10/2021	25/10/2021
									A070101	Failure to Capture	25/10/2021	25/10/2021
									A0709	Device Sensing Problem	25/10/2021	25/10/2021
									A070909	Oversensing	25/10/2021	25/10/2021
									A072201	High Impedance	25/10/2021	25/10/2021
							F1901	Additional Surgery	A0203	Defective Device	25/10/2021	25/10/2021
									A070101	Failure to Capture	25/10/2021	25/10/2021
									A0709	Device Sensing Problem	25/10/2021	25/10/2021
									A070909	Oversensing	25/10/2021	25/10/2021
									A072201	High Impedance	25/10/2021	25/10/2021
6959	2021	All other reportable incidents	35852	Single-chamber implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem	01/12/2021	01/12/2021
									A070504	Premature Discharge of Battery	01/12/2021	01/12/2021
							F08	Hospitalization or Prolonged Hospitalization	A0705	Battery Problem	01/12/2021	01/12/2021
									A070504	Premature Discharge of Battery	01/12/2021	01/12/2021
							F1905	Device Revision or Replacement	A0705	Battery Problem	01/12/2021	01/12/2021
									A070504	Premature Discharge of Battery	01/12/2021	01/12/2021
6960	2022	All other reportable incidents	47263	Cardiac resynchronization therapy implantable pacemaker	CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A0701	Capturing Problem	14/06/2022	14/06/2022
									A090805	Incorrect Measurement	14/06/2022	14/06/2022
							F1901	Additional Surgery	A0701	Capturing Problem	14/06/2022	14/06/2022
									A090805	Incorrect Measurement	14/06/2022	14/06/2022
					CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A0701	Capturing Problem	14/06/2022	14/06/2022
									A090805	Incorrect Measurement	14/06/2022	14/06/2022
							F1901	Additional Surgery	A0701	Capturing Problem	14/06/2022	14/06/2022
									A090805	Incorrect Measurement	14/06/2022	14/06/2022
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A0701	Capturing Problem	14/06/2022	14/06/2022
									A090805	Incorrect Measurement	14/06/2022	14/06/2022
							F1901	Additional Surgery	A0701	Capturing Problem	14/06/2022	14/06/2022
									A090805	Incorrect Measurement	14/06/2022	14/06/2022
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A0701	Capturing Problem	14/06/2022	14/06/2022
									A090805	Incorrect Measurement	14/06/2022	14/06/2022
							F1901	Additional Surgery	A0701	Capturing Problem	14/06/2022	14/06/2022
									A090805	Incorrect Measurement	14/06/2022	14/06/2022
6961	2022	All other reportable incidents	47265	Dual-chamber implantable pacemaker, rate-responsive	CT141	Electrodes/Leads and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT145	Cardiovascular devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT233	Surgical	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT2406	Implantable	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT2407	Active implantable	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT2727	Active implant devices/ancillaries	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT2750	Implantable cardiac electrical stimulators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT2985	Active	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT3001	Implantable metal	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT3002	Implantable synthetic polymer	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT3006	Implantable inorganic material	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT314	Electrically powered	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT321	Long-term surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT449	Generators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT981	Single use	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT983	Surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
6962	2022	All other reportable incidents	17882	Physiologic-monitoring defibrillation system	CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A0710	Failure to Conduct		
							F27	Problem Identified During Non-Clinical Procedure	A0710	Failure to Conduct		
6963	2022	All other reportable incidents	47265	Dual-chamber implantable pacemaker, rate-responsive	CT141	Electrodes/Leads and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem	13/10/2022	13/10/2022
									A070504	Premature Discharge of Battery	13/10/2022	13/10/2022
							F08	Hospitalization or Prolonged Hospitalization	A0705	Battery Problem	13/10/2022	13/10/2022
									A070504	Premature Discharge of Battery	13/10/2022	13/10/2022
							F1905	Device Revision or Replacement	A0705	Battery Problem	13/10/2022	13/10/2022
									A070504	Premature Discharge of Battery	13/10/2022	13/10/2022
					CT145	Cardiovascular devices	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem	13/10/2022	13/10/2022
									A070504	Premature Discharge of Battery	13/10/2022	13/10/2022
							F08	Hospitalization or Prolonged Hospitalization	A0705	Battery Problem	13/10/2022	13/10/2022
									A070504	Premature Discharge of Battery	13/10/2022	13/10/2022
							F1905	Device Revision or Replacement	A0705	Battery Problem	13/10/2022	13/10/2022
									A070504	Premature Discharge of Battery	13/10/2022	13/10/2022

					CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem	13/10/2022	13/10/2022
									A070504	Premature Discharge of Battery	13/10/2022	13/10/2022
							F08	Hospitalization or Prolonged Hospitalization	A0705	Battery Problem	13/10/2022	13/10/2022
									A070504	Premature Discharge of Battery	13/10/2022	13/10/2022
							F1905	Device Revision or Replacement	A0705	Battery Problem	13/10/2022	13/10/2022
									A070504	Premature Discharge of Battery	13/10/2022	13/10/2022
					CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem	13/10/2022	13/10/2022
									A070504	Premature Discharge of Battery	13/10/2022	13/10/2022
							F08	Hospitalization or Prolonged Hospitalization	A0705	Battery Problem	13/10/2022	13/10/2022
									A070504	Premature Discharge of Battery	13/10/2022	13/10/2022
							F1905	Device Revision or Replacement	A0705	Battery Problem	13/10/2022	13/10/2022
									A070504	Premature Discharge of Battery	13/10/2022	13/10/2022
					CT233	Surgical	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem	13/10/2022	13/10/2022
									A070504	Premature Discharge of Battery	13/10/2022	13/10/2022
							F08	Hospitalization or Prolonged Hospitalization	A0705	Battery Problem	13/10/2022	13/10/2022
									A070504	Premature Discharge of Battery	13/10/2022	13/10/2022
							F1905	Device Revision or Replacement	A0705	Battery Problem	13/10/2022	13/10/2022
									A070504	Premature Discharge of Battery	13/10/2022	13/10/2022
					CT2406	Implantable	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem	13/10/2022	13/10/2022
									A070504	Premature Discharge of Battery	13/10/2022	13/10/2022
							F08	Hospitalization or Prolonged Hospitalization	A0705	Battery Problem	13/10/2022	13/10/2022
									A070504	Premature Discharge of Battery	13/10/2022	13/10/2022
							F1905	Device Revision or Replacement	A0705	Battery Problem	13/10/2022	13/10/2022
									A070504	Premature Discharge of Battery	13/10/2022	13/10/2022
					CT2407	Active implantable	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem	13/10/2022	13/10/2022
									A070504	Premature Discharge of Battery	13/10/2022	13/10/2022
							F08	Hospitalization or Prolonged Hospitalization	A0705	Battery Problem	13/10/2022	13/10/2022
									A070504	Premature Discharge of Battery	13/10/2022	13/10/2022
							F1905	Device Revision or Replacement	A0705	Battery Problem	13/10/2022	13/10/2022
									A070504	Premature Discharge of Battery	13/10/2022	13/10/2022
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem	13/10/2022	13/10/2022
									A070504	Premature Discharge of Battery	13/10/2022	13/10/2022
							F08	Hospitalization or Prolonged Hospitalization	A0705	Battery Problem	13/10/2022	13/10/2022
									A070504	Premature Discharge of Battery	13/10/2022	13/10/2022
							F1905	Device Revision or Replacement	A0705	Battery Problem	13/10/2022	13/10/2022
									A070504	Premature Discharge of Battery	13/10/2022	13/10/2022
					CT2727	Active implant devices/ancillaries	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem	13/10/2022	13/10/2022
									A070504	Premature Discharge of Battery	13/10/2022	13/10/2022
							F08	Hospitalization or Prolonged Hospitalization	A0705	Battery Problem	13/10/2022	13/10/2022
									A070504	Premature Discharge of Battery	13/10/2022	13/10/2022
							F1905	Device Revision or Replacement	A0705	Battery Problem	13/10/2022	13/10/2022
									A070504	Premature Discharge of Battery	13/10/2022	13/10/2022
					CT2750	Implantable cardiac electrical stimulators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem	13/10/2022	13/10/2022
									A070504	Premature Discharge of Battery	13/10/2022	13/10/2022
							F08	Hospitalization or Prolonged Hospitalization	A0705	Battery Problem	13/10/2022	13/10/2022
									A070504	Premature Discharge of Battery	13/10/2022	13/10/2022
							F1905	Device Revision or Replacement	A0705	Battery Problem	13/10/2022	13/10/2022
									A070504	Premature Discharge of Battery	13/10/2022	13/10/2022
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem	13/10/2022	13/10/2022
									A070504	Premature Discharge of Battery	13/10/2022	13/10/2022
							F08	Hospitalization or Prolonged Hospitalization	A0705	Battery Problem	13/10/2022	13/10/2022
									A070504	Premature Discharge of Battery	13/10/2022	13/10/2022
							F1905	Device Revision or Replacement	A0705	Battery Problem	13/10/2022	13/10/2022
									A070504	Premature Discharge of Battery	13/10/2022	13/10/2022
					CT2985	Active	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem	13/10/2022	13/10/2022
									A070504	Premature Discharge of Battery	13/10/2022	13/10/2022
							F08	Hospitalization or Prolonged Hospitalization	A0705	Battery Problem	13/10/2022	13/10/2022
									A070504	Premature Discharge of Battery	13/10/2022	13/10/2022
							F1905	Device Revision or Replacement	A0705	Battery Problem	13/10/2022	13/10/2022
									A070504	Premature Discharge of Battery	13/10/2022	13/10/2022
					CT3001	Implantable metal	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem	13/10/2022	13/10/2022
									A070504	Premature Discharge of Battery	13/10/2022	13/10/2022
							F08	Hospitalization or Prolonged Hospitalization	A0705	Battery Problem	13/10/2022	13/10/2022
									A070504	Premature Discharge of Battery	13/10/2022	13/10/2022
							F1905	Device Revision or Replacement	A0705	Battery Problem	13/10/2022	13/10/2022
									A070504	Premature Discharge of Battery	13/10/2022	13/10/2022
					CT3002	Implantable synthetic polymer	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem	13/10/2022	13/10/2022
									A070504	Premature Discharge of Battery	13/10/2022	13/10/2022
							F08	Hospitalization or Prolonged Hospitalization	A0705	Battery Problem	13/10/2022	13/10/2022
									A070504	Premature Discharge of Battery	13/10/2022	13/10/2022
							F1905	Device Revision or Replacement	A0705	Battery Problem	13/10/2022	13/10/2022
									A070504	Premature Discharge of Battery	13/10/2022	13/10/2022
					CT3006	Implantable inorganic material	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem	13/10/2022	13/10/2022
									A070504	Premature Discharge of Battery	13/10/2022	13/10/2022
							F08	Hospitalization or Prolonged Hospitalization	A0705	Battery Problem	13/10/2022	13/10/2022
									A070504	Premature Discharge of Battery	13/10/2022	13/10/2022
							F1905	Device Revision or Replacement	A0705	Battery Problem	13/10/2022	13/10/2022
									A070504	Premature Discharge of Battery	13/10/2022	13/10/2022
					CT314	Electrically powered	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem	13/10/2022	13/10/2022
									A070504	Premature Discharge of Battery	13/10/2022	13/10/2022
							F08	Hospitalization or Prolonged Hospitalization	A0705	Battery Problem	13/10/2022	13/10/2022
									A070504	Premature Discharge of Battery	13/10/2022	13/10/2022
							F1905	Device Revision or Replacement	A0705	Battery Problem	13/10/2022	13/10/2022
									A070504	Premature Discharge of Battery	13/10/2022	13/10/2022
					CT321	Long-term surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem	13/10/2022	13/10/2022
									A070504	Premature Discharge of Battery	13/10/2022	13/10/2022
							F08	Hospitalization or Prolonged Hospitalization	A0705	Battery Problem	13/10/2022	13/10/2022
									A070504	Premature Discharge of Battery	13/10/2022	13/10/2022

							F1905	Device Revision or Replacement	A0705	Battery Problem	13/10/2022	13/10/2022
									A070504	Premature Discharge of Battery	13/10/2022	13/10/2022
					CT449	Generators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem	13/10/2022	13/10/2022
									A070504	Premature Discharge of Battery	13/10/2022	13/10/2022
							F08	Hospitalization or Prolonged Hospitalization	A0705	Battery Problem	13/10/2022	13/10/2022
									A070504	Premature Discharge of Battery	13/10/2022	13/10/2022
							F1905	Device Revision or Replacement	A0705	Battery Problem	13/10/2022	13/10/2022
									A070504	Premature Discharge of Battery	13/10/2022	13/10/2022
					CT981	Single use	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem	13/10/2022	13/10/2022
									A070504	Premature Discharge of Battery	13/10/2022	13/10/2022
							F08	Hospitalization or Prolonged Hospitalization	A0705	Battery Problem	13/10/2022	13/10/2022
									A070504	Premature Discharge of Battery	13/10/2022	13/10/2022
							F1905	Device Revision or Replacement	A0705	Battery Problem	13/10/2022	13/10/2022
									A070504	Premature Discharge of Battery	13/10/2022	13/10/2022
					CT983	Surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem	13/10/2022	13/10/2022
									A070504	Premature Discharge of Battery	13/10/2022	13/10/2022
							F08	Hospitalization or Prolonged Hospitalization	A0705	Battery Problem	13/10/2022	13/10/2022
									A070504	Premature Discharge of Battery	13/10/2022	13/10/2022
							F1905	Device Revision or Replacement	A0705	Battery Problem	13/10/2022	13/10/2022
									A070504	Premature Discharge of Battery	13/10/2022	13/10/2022
6964	2021	All other reportable incidents	47265	Dual-chamber implantable pacemaker, rate-responsive	CT141	Electrodes/Leads and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	03/11/2021	03/11/2021
									A0711	Interrogation Problem	03/11/2021	03/11/2021
							F1901	Additional Surgery	A070504	Premature Discharge of Battery	03/11/2021	03/11/2021
									A0711	Interrogation Problem	03/11/2021	03/11/2021
					CT145	Cardiovascular devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	03/11/2021	03/11/2021
									A0711	Interrogation Problem	03/11/2021	03/11/2021
							F1901	Additional Surgery	A070504	Premature Discharge of Battery	03/11/2021	03/11/2021
									A0711	Interrogation Problem	03/11/2021	03/11/2021
					CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	03/11/2021	03/11/2021
									A0711	Interrogation Problem	03/11/2021	03/11/2021
							F1901	Additional Surgery	A070504	Premature Discharge of Battery	03/11/2021	03/11/2021
									A0711	Interrogation Problem	03/11/2021	03/11/2021
					CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	03/11/2021	03/11/2021
									A0711	Interrogation Problem	03/11/2021	03/11/2021
							F1901	Additional Surgery	A070504	Premature Discharge of Battery	03/11/2021	03/11/2021
									A0711	Interrogation Problem	03/11/2021	03/11/2021
					CT233	Surgical	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	03/11/2021	03/11/2021
									A0711	Interrogation Problem	03/11/2021	03/11/2021
							F1901	Additional Surgery	A070504	Premature Discharge of Battery	03/11/2021	03/11/2021
									A0711	Interrogation Problem	03/11/2021	03/11/2021
					CT2406	Implantable	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	03/11/2021	03/11/2021
									A0711	Interrogation Problem	03/11/2021	03/11/2021
							F1901	Additional Surgery	A070504	Premature Discharge of Battery	03/11/2021	03/11/2021
									A0711	Interrogation Problem	03/11/2021	03/11/2021
					CT2407	Active implantable	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	03/11/2021	03/11/2021
									A0711	Interrogation Problem	03/11/2021	03/11/2021
							F1901	Additional Surgery	A070504	Premature Discharge of Battery	03/11/2021	03/11/2021
									A0711	Interrogation Problem	03/11/2021	03/11/2021
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	03/11/2021	03/11/2021
									A0711	Interrogation Problem	03/11/2021	03/11/2021
							F1901	Additional Surgery	A070504	Premature Discharge of Battery	03/11/2021	03/11/2021
									A0711	Interrogation Problem	03/11/2021	03/11/2021
					CT2727	Active implant devices/ancillaries	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	03/11/2021	03/11/2021
									A0711	Interrogation Problem	03/11/2021	03/11/2021
							F1901	Additional Surgery	A070504	Premature Discharge of Battery	03/11/2021	03/11/2021
									A0711	Interrogation Problem	03/11/2021	03/11/2021
					CT2750	Implantable cardiac electrical stimulators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	03/11/2021	03/11/2021
									A0711	Interrogation Problem	03/11/2021	03/11/2021
							F1901	Additional Surgery	A070504	Premature Discharge of Battery	03/11/2021	03/11/2021
									A0711	Interrogation Problem	03/11/2021	03/11/2021
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	03/11/2021	03/11/2021
									A0711	Interrogation Problem	03/11/2021	03/11/2021
							F1901	Additional Surgery	A070504	Premature Discharge of Battery	03/11/2021	03/11/2021
									A0711	Interrogation Problem	03/11/2021	03/11/2021
					CT2985	Active	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	03/11/2021	03/11/2021
									A0711	Interrogation Problem	03/11/2021	03/11/2021
							F1901	Additional Surgery	A070504	Premature Discharge of Battery	03/11/2021	03/11/2021
									A0711	Interrogation Problem	03/11/2021	03/11/2021
					CT3001	Implantable metal	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	03/11/2021	03/11/2021
									A0711	Interrogation Problem	03/11/2021	03/11/2021
							F1901	Additional Surgery	A070504	Premature Discharge of Battery	03/11/2021	03/11/2021
									A0711	Interrogation Problem	03/11/2021	03/11/2021
					CT3002	Implantable synthetic polymer	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	03/11/2021	03/11/2021
									A0711	Interrogation Problem	03/11/2021	03/11/2021
							F1901	Additional Surgery	A070504	Premature Discharge of Battery	03/11/2021	03/11/2021
									A0711	Interrogation Problem	03/11/2021	03/11/2021
					CT3006	Implantable inorganic material	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	03/11/2021	03/11/2021
									A0711	Interrogation Problem	03/11/2021	03/11/2021
							F1901	Additional Surgery	A070504	Premature Discharge of Battery	03/11/2021	03/11/2021
									A0711	Interrogation Problem	03/11/2021	03/11/2021
					CT314	Electrically powered	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	03/11/2021	03/11/2021
									A0711	Interrogation Problem	03/11/2021	03/11/2021
							F1901	Additional Surgery	A070504	Premature Discharge of Battery	03/11/2021	03/11/2021
									A0711	Interrogation Problem	03/11/2021	03/11/2021
					CT321	Long-term surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	03/11/2021	03/11/2021
									A0711	Interrogation Problem	03/11/2021	03/11/2021
							F1901	Additional Surgery	A070504	Premature Discharge of Battery	03/11/2021	03/11/2021
									A0711	Interrogation Problem	03/11/2021	03/11/2021

					CT449	Generators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	03/11/2021	03/11/2021
									A0711	Interrogation Problem	03/11/2021	03/11/2021
							F1901	Additional Surgery	A070504	Premature Discharge of Battery	03/11/2021	03/11/2021
									A0711	Interrogation Problem	03/11/2021	03/11/2021
					CT981	Single use	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	03/11/2021	03/11/2021
									A0711	Interrogation Problem	03/11/2021	03/11/2021
							F1901	Additional Surgery	A070504	Premature Discharge of Battery	03/11/2021	03/11/2021
									A0711	Interrogation Problem	03/11/2021	03/11/2021
					CT983	Surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	03/11/2021	03/11/2021
									A0711	Interrogation Problem	03/11/2021	03/11/2021
							F1901	Additional Surgery	A070504	Premature Discharge of Battery	03/11/2021	03/11/2021
									A0711	Interrogation Problem	03/11/2021	03/11/2021
6965	2022	All other reportable incidents	47263	Cardiac resynchronization therapy implantable pacemaker	CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A0908	Incorrect, Inadequate or Imprecise Result or Readings	30/05/2022	30/05/2022
							F08	Hospitalization or Prolonged Hospitalization	A0908	Incorrect, Inadequate or Imprecise Result or Readings	30/05/2022	30/05/2022
							F1905	Device Revision or Replacement	A0908	Incorrect, Inadequate or Imprecise Result or Readings	30/05/2022	30/05/2022
					CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A0908	Incorrect, Inadequate or Imprecise Result or Readings	30/05/2022	30/05/2022
							F08	Hospitalization or Prolonged Hospitalization	A0908	Incorrect, Inadequate or Imprecise Result or Readings	30/05/2022	30/05/2022
							F1905	Device Revision or Replacement	A0908	Incorrect, Inadequate or Imprecise Result or Readings	30/05/2022	30/05/2022
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A0908	Incorrect, Inadequate or Imprecise Result or Readings	30/05/2022	30/05/2022
							F08	Hospitalization or Prolonged Hospitalization	A0908	Incorrect, Inadequate or Imprecise Result or Readings	30/05/2022	30/05/2022
							F1905	Device Revision or Replacement	A0908	Incorrect, Inadequate or Imprecise Result or Readings	30/05/2022	30/05/2022
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A0908	Incorrect, Inadequate or Imprecise Result or Readings	30/05/2022	30/05/2022
							F08	Hospitalization or Prolonged Hospitalization	A0908	Incorrect, Inadequate or Imprecise Result or Readings	30/05/2022	30/05/2022
							F1905	Device Revision or Replacement	A0908	Incorrect, Inadequate or Imprecise Result or Readings	30/05/2022	30/05/2022
6966	2022	All other reportable incidents	47265	Dual-chamber implantable pacemaker, rate-responsive	CT141	Electrodes/Leads and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem	26/08/2022	26/08/2022
									A070504	Premature Discharge of Battery	26/08/2022	26/08/2022
							F08	Hospitalization or Prolonged Hospitalization	A0705	Battery Problem	26/08/2022	26/08/2022
									A070504	Premature Discharge of Battery	26/08/2022	26/08/2022
							F1903	Device Explantation	A0705	Battery Problem	26/08/2022	26/08/2022
									A070504	Premature Discharge of Battery	26/08/2022	26/08/2022
					CT145	Cardiovascular devices	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem	26/08/2022	26/08/2022
									A070504	Premature Discharge of Battery	26/08/2022	26/08/2022
							F08	Hospitalization or Prolonged Hospitalization	A0705	Battery Problem	26/08/2022	26/08/2022
									A070504	Premature Discharge of Battery	26/08/2022	26/08/2022
							F1903	Device Explantation	A0705	Battery Problem	26/08/2022	26/08/2022
									A070504	Premature Discharge of Battery	26/08/2022	26/08/2022
					CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem	26/08/2022	26/08/2022
									A070504	Premature Discharge of Battery	26/08/2022	26/08/2022
							F08	Hospitalization or Prolonged Hospitalization	A0705	Battery Problem	26/08/2022	26/08/2022
									A070504	Premature Discharge of Battery	26/08/2022	26/08/2022
							F1903	Device Explantation	A0705	Battery Problem	26/08/2022	26/08/2022
									A070504	Premature Discharge of Battery	26/08/2022	26/08/2022
					CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem	26/08/2022	26/08/2022
									A070504	Premature Discharge of Battery	26/08/2022	26/08/2022
							F08	Hospitalization or Prolonged Hospitalization	A0705	Battery Problem	26/08/2022	26/08/2022
									A070504	Premature Discharge of Battery	26/08/2022	26/08/2022
							F1903	Device Explantation	A0705	Battery Problem	26/08/2022	26/08/2022
									A070504	Premature Discharge of Battery	26/08/2022	26/08/2022
					CT233	Surgical	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem	26/08/2022	26/08/2022
									A070504	Premature Discharge of Battery	26/08/2022	26/08/2022
							F08	Hospitalization or Prolonged Hospitalization	A0705	Battery Problem	26/08/2022	26/08/2022
									A070504	Premature Discharge of Battery	26/08/2022	26/08/2022
							F1903	Device Explantation	A0705	Battery Problem	26/08/2022	26/08/2022
									A070504	Premature Discharge of Battery	26/08/2022	26/08/2022
					CT2406	Implantable	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem	26/08/2022	26/08/2022
									A070504	Premature Discharge of Battery	26/08/2022	26/08/2022
							F08	Hospitalization or Prolonged Hospitalization	A0705	Battery Problem	26/08/2022	26/08/2022
									A070504	Premature Discharge of Battery	26/08/2022	26/08/2022
							F1903	Device Explantation	A0705	Battery Problem	26/08/2022	26/08/2022
									A070504	Premature Discharge of Battery	26/08/2022	26/08/2022
					CT2407	Active implantable	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem	26/08/2022	26/08/2022
									A070504	Premature Discharge of Battery	26/08/2022	26/08/2022
							F08	Hospitalization or Prolonged Hospitalization	A0705	Battery Problem	26/08/2022	26/08/2022
									A070504	Premature Discharge of Battery	26/08/2022	26/08/2022
							F1903	Device Explantation	A0705	Battery Problem	26/08/2022	26/08/2022
									A070504	Premature Discharge of Battery	26/08/2022	26/08/2022
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem	26/08/2022	26/08/2022
									A070504	Premature Discharge of Battery	26/08/2022	26/08/2022
							F08	Hospitalization or Prolonged Hospitalization	A0705	Battery Problem	26/08/2022	26/08/2022
									A070504	Premature Discharge of Battery	26/08/2022	26/08/2022
							F1903	Device Explantation	A0705	Battery Problem	26/08/2022	26/08/2022
									A070504	Premature Discharge of Battery	26/08/2022	26/08/2022
					CT2727	Active implant devices/ancillaries	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem	26/08/2022	26/08/2022
									A070504	Premature Discharge of Battery	26/08/2022	26/08/2022
							F08	Hospitalization or Prolonged Hospitalization	A0705	Battery Problem	26/08/2022	26/08/2022
									A070504	Premature Discharge of Battery	26/08/2022	26/08/2022
							F1903	Device Explantation	A0705	Battery Problem	26/08/2022	26/08/2022
									A070504	Premature Discharge of Battery	26/08/2022	26/08/2022
					CT2750	Implantable cardiac electrical stimulators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem	26/08/2022	26/08/2022

									A070504	Premature Discharge of Battery	26/08/2022	26/08/2022
							F08	Hospitalization or Prolonged Hospitalization	A0705	Battery Problem	26/08/2022	26/08/2022
									A070504	Premature Discharge of Battery	26/08/2022	26/08/2022
							F1903	Device Explantation	A0705	Battery Problem	26/08/2022	26/08/2022
									A070504	Premature Discharge of Battery	26/08/2022	26/08/2022
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem	26/08/2022	26/08/2022
									A070504	Premature Discharge of Battery	26/08/2022	26/08/2022
							F08	Hospitalization or Prolonged Hospitalization	A0705	Battery Problem	26/08/2022	26/08/2022
									A070504	Premature Discharge of Battery	26/08/2022	26/08/2022
							F1903	Device Explantation	A0705	Battery Problem	26/08/2022	26/08/2022
									A070504	Premature Discharge of Battery	26/08/2022	26/08/2022
					CT2985	Active	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem	26/08/2022	26/08/2022
									A070504	Premature Discharge of Battery	26/08/2022	26/08/2022
							F08	Hospitalization or Prolonged Hospitalization	A0705	Battery Problem	26/08/2022	26/08/2022
									A070504	Premature Discharge of Battery	26/08/2022	26/08/2022
							F1903	Device Explantation	A0705	Battery Problem	26/08/2022	26/08/2022
									A070504	Premature Discharge of Battery	26/08/2022	26/08/2022
					CT3001	Implantable metal	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem	26/08/2022	26/08/2022
									A070504	Premature Discharge of Battery	26/08/2022	26/08/2022
							F08	Hospitalization or Prolonged Hospitalization	A0705	Battery Problem	26/08/2022	26/08/2022
									A070504	Premature Discharge of Battery	26/08/2022	26/08/2022
							F1903	Device Explantation	A0705	Battery Problem	26/08/2022	26/08/2022
									A070504	Premature Discharge of Battery	26/08/2022	26/08/2022
					CT3002	Implantable synthetic polymer	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem	26/08/2022	26/08/2022
									A070504	Premature Discharge of Battery	26/08/2022	26/08/2022
							F08	Hospitalization or Prolonged Hospitalization	A0705	Battery Problem	26/08/2022	26/08/2022
									A070504	Premature Discharge of Battery	26/08/2022	26/08/2022
							F1903	Device Explantation	A0705	Battery Problem	26/08/2022	26/08/2022
									A070504	Premature Discharge of Battery	26/08/2022	26/08/2022
					CT3006	Implantable inorganic material	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem	26/08/2022	26/08/2022
									A070504	Premature Discharge of Battery	26/08/2022	26/08/2022
							F08	Hospitalization or Prolonged Hospitalization	A0705	Battery Problem	26/08/2022	26/08/2022
									A070504	Premature Discharge of Battery	26/08/2022	26/08/2022
							F1903	Device Explantation	A0705	Battery Problem	26/08/2022	26/08/2022
									A070504	Premature Discharge of Battery	26/08/2022	26/08/2022
					CT314	Electrically powered	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem	26/08/2022	26/08/2022
									A070504	Premature Discharge of Battery	26/08/2022	26/08/2022
							F08	Hospitalization or Prolonged Hospitalization	A0705	Battery Problem	26/08/2022	26/08/2022
									A070504	Premature Discharge of Battery	26/08/2022	26/08/2022
							F1903	Device Explantation	A0705	Battery Problem	26/08/2022	26/08/2022
									A070504	Premature Discharge of Battery	26/08/2022	26/08/2022
					CT321	Long-term surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem	26/08/2022	26/08/2022
									A070504	Premature Discharge of Battery	26/08/2022	26/08/2022
							F08	Hospitalization or Prolonged Hospitalization	A0705	Battery Problem	26/08/2022	26/08/2022
									A070504	Premature Discharge of Battery	26/08/2022	26/08/2022
							F1903	Device Explantation	A0705	Battery Problem	26/08/2022	26/08/2022
									A070504	Premature Discharge of Battery	26/08/2022	26/08/2022
					CT449	Generators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem	26/08/2022	26/08/2022
									A070504	Premature Discharge of Battery	26/08/2022	26/08/2022
							F08	Hospitalization or Prolonged Hospitalization	A0705	Battery Problem	26/08/2022	26/08/2022
									A070504	Premature Discharge of Battery	26/08/2022	26/08/2022
							F1903	Device Explantation	A0705	Battery Problem	26/08/2022	26/08/2022
									A070504	Premature Discharge of Battery	26/08/2022	26/08/2022
					CT981	Single use	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem	26/08/2022	26/08/2022
									A070504	Premature Discharge of Battery	26/08/2022	26/08/2022
							F08	Hospitalization or Prolonged Hospitalization	A0705	Battery Problem	26/08/2022	26/08/2022
									A070504	Premature Discharge of Battery	26/08/2022	26/08/2022
							F1903	Device Explantation	A0705	Battery Problem	26/08/2022	26/08/2022
									A070504	Premature Discharge of Battery	26/08/2022	26/08/2022
					CT983	Surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem	26/08/2022	26/08/2022
									A070504	Premature Discharge of Battery	26/08/2022	26/08/2022
							F08	Hospitalization or Prolonged Hospitalization	A0705	Battery Problem	26/08/2022	26/08/2022
									A070504	Premature Discharge of Battery	26/08/2022	26/08/2022
							F1903	Device Explantation	A0705	Battery Problem	26/08/2022	26/08/2022
									A070504	Premature Discharge of Battery	26/08/2022	26/08/2022
6967	2021	All other reportable incidents	36241	Pacemaker/defibrillator lead	CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A05	Mechanical Problem	26/10/2021	26/10/2021
									A0709	Device Sensing Problem	26/10/2021	26/10/2021
							F1908	Prolonged Surgery	A05	Mechanical Problem	26/10/2021	26/10/2021
									A0709	Device Sensing Problem	26/10/2021	26/10/2021
6968	2021	All other reportable incidents	47265	Dual-chamber implantable pacemaker, rate-responsive	CT141	Electrodes/Leads and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	17/11/2021	17/11/2021
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	17/11/2021	17/11/2021
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	17/11/2021	17/11/2021
					CT145	Cardiovascular devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	17/11/2021	17/11/2021
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	17/11/2021	17/11/2021
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	17/11/2021	17/11/2021
					CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	17/11/2021	17/11/2021
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	17/11/2021	17/11/2021
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	17/11/2021	17/11/2021
					CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	17/11/2021	17/11/2021
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	17/11/2021	17/11/2021
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	17/11/2021	17/11/2021
					CT233	Surgical	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	17/11/2021	17/11/2021
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	17/11/2021	17/11/2021
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	17/11/2021	17/11/2021
					CT2406	Implantable	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	17/11/2021	17/11/2021
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	17/11/2021	17/11/2021
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	17/11/2021	17/11/2021
					CT2407	Active implantable	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	17/11/2021	17/11/2021

							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	17/11/2021	17/11/2021
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	17/11/2021	17/11/2021
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	17/11/2021	17/11/2021
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	17/11/2021	17/11/2021
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	17/11/2021	17/11/2021
					CT2727	Active implant devices/ancillaries	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	17/11/2021	17/11/2021
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	17/11/2021	17/11/2021
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	17/11/2021	17/11/2021
					CT2750	Implantable cardiac electrical stimulators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	17/11/2021	17/11/2021
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	17/11/2021	17/11/2021
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	17/11/2021	17/11/2021
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	17/11/2021	17/11/2021
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	17/11/2021	17/11/2021
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	17/11/2021	17/11/2021
					CT2985	Active	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	17/11/2021	17/11/2021
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	17/11/2021	17/11/2021
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	17/11/2021	17/11/2021
					CT3001	Implantable metal	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	17/11/2021	17/11/2021
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	17/11/2021	17/11/2021
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	17/11/2021	17/11/2021
					CT3002	Implantable synthetic polymer	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	17/11/2021	17/11/2021
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	17/11/2021	17/11/2021
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	17/11/2021	17/11/2021
					CT3006	Implantable inorganic material	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	17/11/2021	17/11/2021
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	17/11/2021	17/11/2021
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	17/11/2021	17/11/2021
					CT314	Electrically powered	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	17/11/2021	17/11/2021
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	17/11/2021	17/11/2021
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	17/11/2021	17/11/2021
					CT321	Long-term surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	17/11/2021	17/11/2021
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	17/11/2021	17/11/2021
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	17/11/2021	17/11/2021
					CT449	Generators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	17/11/2021	17/11/2021
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	17/11/2021	17/11/2021
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	17/11/2021	17/11/2021
					CT981	Single use	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	17/11/2021	17/11/2021
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	17/11/2021	17/11/2021
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	17/11/2021	17/11/2021
					CT983	Surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	17/11/2021	17/11/2021
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	17/11/2021	17/11/2021
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	17/11/2021	17/11/2021
6969	2022	All other reportable incidents	47267	Single-chamber implantable pacemaker, rate-responsive	CT141	Electrodes/Leads and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem	01/09/2022	01/09/2022
									A070504	Premature Discharge of Battery	01/09/2022	01/09/2022
							F08	Hospitalization or Prolonged Hospitalization	A0705	Battery Problem	01/09/2022	01/09/2022
									A070504	Premature Discharge of Battery	01/09/2022	01/09/2022
							F1905	Device Revision or Replacement	A0705	Battery Problem	01/09/2022	01/09/2022
									A070504	Premature Discharge of Battery	01/09/2022	01/09/2022
					CT145	Cardiovascular devices	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem	01/09/2022	01/09/2022
									A070504	Premature Discharge of Battery	01/09/2022	01/09/2022
							F08	Hospitalization or Prolonged Hospitalization	A0705	Battery Problem	01/09/2022	01/09/2022
									A070504	Premature Discharge of Battery	01/09/2022	01/09/2022
							F1905	Device Revision or Replacement	A0705	Battery Problem	01/09/2022	01/09/2022
									A070504	Premature Discharge of Battery	01/09/2022	01/09/2022
					CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem	01/09/2022	01/09/2022
									A070504	Premature Discharge of Battery	01/09/2022	01/09/2022
							F08	Hospitalization or Prolonged Hospitalization	A0705	Battery Problem	01/09/2022	01/09/2022
									A070504	Premature Discharge of Battery	01/09/2022	01/09/2022
							F1905	Device Revision or Replacement	A0705	Battery Problem	01/09/2022	01/09/2022
									A070504	Premature Discharge of Battery	01/09/2022	01/09/2022
					CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem	01/09/2022	01/09/2022
									A070504	Premature Discharge of Battery	01/09/2022	01/09/2022
							F08	Hospitalization or Prolonged Hospitalization	A0705	Battery Problem	01/09/2022	01/09/2022
									A070504	Premature Discharge of Battery	01/09/2022	01/09/2022
							F1905	Device Revision or Replacement	A0705	Battery Problem	01/09/2022	01/09/2022
									A070504	Premature Discharge of Battery	01/09/2022	01/09/2022
					CT233	Surgical	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem	01/09/2022	01/09/2022
									A070504	Premature Discharge of Battery	01/09/2022	01/09/2022
							F08	Hospitalization or Prolonged Hospitalization	A0705	Battery Problem	01/09/2022	01/09/2022
									A070504	Premature Discharge of Battery	01/09/2022	01/09/2022
							F1905	Device Revision or Replacement	A0705	Battery Problem	01/09/2022	01/09/2022
									A070504	Premature Discharge of Battery	01/09/2022	01/09/2022
					CT2406	Implantable	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem	01/09/2022	01/09/2022
									A070504	Premature Discharge of Battery	01/09/2022	01/09/2022
							F08	Hospitalization or Prolonged Hospitalization	A0705	Battery Problem	01/09/2022	01/09/2022
									A070504	Premature Discharge of Battery	01/09/2022	01/09/2022
							F1905	Device Revision or Replacement	A0705	Battery Problem	01/09/2022	01/09/2022
									A070504	Premature Discharge of Battery	01/09/2022	01/09/2022
					CT2407	Active implantable	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem	01/09/2022	01/09/2022
									A070504	Premature Discharge of Battery	01/09/2022	01/09/2022
							F08	Hospitalization or Prolonged Hospitalization	A0705	Battery Problem	01/09/2022	01/09/2022
									A070504	Premature Discharge of Battery	01/09/2022	01/09/2022
							F1905	Device Revision or Replacement	A0705	Battery Problem	01/09/2022	01/09/2022
									A070504	Premature Discharge of Battery	01/09/2022	01/09/2022
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem	01/09/2022	01/09/2022
									A070504	Premature Discharge of Battery	01/09/2022	01/09/2022
							F08	Hospitalization or Prolonged Hospitalization	A0705	Battery Problem	01/09/2022	01/09/2022
									A070504	Premature Discharge of Battery	01/09/2022	01/09/2022
							F1905	Device Revision or Replacement	A0705	Battery Problem	01/09/2022	01/09/2022

									A070504	Premature Discharge of Battery	01/09/2022	01/09/2022
					CT2727	Active implant devices/ancillaries	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem	01/09/2022	01/09/2022
									A070504	Premature Discharge of Battery	01/09/2022	01/09/2022
							F08	Hospitalization or Prolonged Hospitalization	A0705	Battery Problem	01/09/2022	01/09/2022
									A070504	Premature Discharge of Battery	01/09/2022	01/09/2022
							F1905	Device Revision or Replacement	A0705	Battery Problem	01/09/2022	01/09/2022
									A070504	Premature Discharge of Battery	01/09/2022	01/09/2022
					CT2750	Implantable cardiac electrical stimulators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem	01/09/2022	01/09/2022
									A070504	Premature Discharge of Battery	01/09/2022	01/09/2022
							F08	Hospitalization or Prolonged Hospitalization	A0705	Battery Problem	01/09/2022	01/09/2022
									A070504	Premature Discharge of Battery	01/09/2022	01/09/2022
							F1905	Device Revision or Replacement	A0705	Battery Problem	01/09/2022	01/09/2022
									A070504	Premature Discharge of Battery	01/09/2022	01/09/2022
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem	01/09/2022	01/09/2022
									A070504	Premature Discharge of Battery	01/09/2022	01/09/2022
							F08	Hospitalization or Prolonged Hospitalization	A0705	Battery Problem	01/09/2022	01/09/2022
									A070504	Premature Discharge of Battery	01/09/2022	01/09/2022
							F1905	Device Revision or Replacement	A0705	Battery Problem	01/09/2022	01/09/2022
									A070504	Premature Discharge of Battery	01/09/2022	01/09/2022
					CT2985	Active	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem	01/09/2022	01/09/2022
									A070504	Premature Discharge of Battery	01/09/2022	01/09/2022
							F08	Hospitalization or Prolonged Hospitalization	A0705	Battery Problem	01/09/2022	01/09/2022
									A070504	Premature Discharge of Battery	01/09/2022	01/09/2022
							F1905	Device Revision or Replacement	A0705	Battery Problem	01/09/2022	01/09/2022
									A070504	Premature Discharge of Battery	01/09/2022	01/09/2022
					CT3001	Implantable metal	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem	01/09/2022	01/09/2022
									A070504	Premature Discharge of Battery	01/09/2022	01/09/2022
							F08	Hospitalization or Prolonged Hospitalization	A0705	Battery Problem	01/09/2022	01/09/2022
									A070504	Premature Discharge of Battery	01/09/2022	01/09/2022
							F1905	Device Revision or Replacement	A0705	Battery Problem	01/09/2022	01/09/2022
									A070504	Premature Discharge of Battery	01/09/2022	01/09/2022
					CT3002	Implantable synthetic polymer	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem	01/09/2022	01/09/2022
									A070504	Premature Discharge of Battery	01/09/2022	01/09/2022
							F08	Hospitalization or Prolonged Hospitalization	A0705	Battery Problem	01/09/2022	01/09/2022
									A070504	Premature Discharge of Battery	01/09/2022	01/09/2022
							F1905	Device Revision or Replacement	A0705	Battery Problem	01/09/2022	01/09/2022
									A070504	Premature Discharge of Battery	01/09/2022	01/09/2022
					CT3006	Implantable inorganic material	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem	01/09/2022	01/09/2022
									A070504	Premature Discharge of Battery	01/09/2022	01/09/2022
							F08	Hospitalization or Prolonged Hospitalization	A0705	Battery Problem	01/09/2022	01/09/2022
									A070504	Premature Discharge of Battery	01/09/2022	01/09/2022
							F1905	Device Revision or Replacement	A0705	Battery Problem	01/09/2022	01/09/2022
									A070504	Premature Discharge of Battery	01/09/2022	01/09/2022
					CT314	Electrically powered	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem	01/09/2022	01/09/2022
									A070504	Premature Discharge of Battery	01/09/2022	01/09/2022
							F08	Hospitalization or Prolonged Hospitalization	A0705	Battery Problem	01/09/2022	01/09/2022
									A070504	Premature Discharge of Battery	01/09/2022	01/09/2022
							F1905	Device Revision or Replacement	A0705	Battery Problem	01/09/2022	01/09/2022
									A070504	Premature Discharge of Battery	01/09/2022	01/09/2022
					CT321	Long-term surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem	01/09/2022	01/09/2022
									A070504	Premature Discharge of Battery	01/09/2022	01/09/2022
							F08	Hospitalization or Prolonged Hospitalization	A0705	Battery Problem	01/09/2022	01/09/2022
									A070504	Premature Discharge of Battery	01/09/2022	01/09/2022
							F1905	Device Revision or Replacement	A0705	Battery Problem	01/09/2022	01/09/2022
									A070504	Premature Discharge of Battery	01/09/2022	01/09/2022
					CT449	Generators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem	01/09/2022	01/09/2022
									A070504	Premature Discharge of Battery	01/09/2022	01/09/2022
							F08	Hospitalization or Prolonged Hospitalization	A0705	Battery Problem	01/09/2022	01/09/2022
									A070504	Premature Discharge of Battery	01/09/2022	01/09/2022
							F1905	Device Revision or Replacement	A0705	Battery Problem	01/09/2022	01/09/2022
									A070504	Premature Discharge of Battery	01/09/2022	01/09/2022
					CT981	Single use	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem	01/09/2022	01/09/2022
									A070504	Premature Discharge of Battery	01/09/2022	01/09/2022
							F08	Hospitalization or Prolonged Hospitalization	A0705	Battery Problem	01/09/2022	01/09/2022
									A070504	Premature Discharge of Battery	01/09/2022	01/09/2022
							F1905	Device Revision or Replacement	A0705	Battery Problem	01/09/2022	01/09/2022
									A070504	Premature Discharge of Battery	01/09/2022	01/09/2022
					CT983	Surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem	01/09/2022	01/09/2022
									A070504	Premature Discharge of Battery	01/09/2022	01/09/2022
							F08	Hospitalization or Prolonged Hospitalization	A0705	Battery Problem	01/09/2022	01/09/2022
									A070504	Premature Discharge of Battery	01/09/2022	01/09/2022
							F1905	Device Revision or Replacement	A0705	Battery Problem	01/09/2022	01/09/2022
									A070504	Premature Discharge of Battery	01/09/2022	01/09/2022
6970	2022	All other reportable incidents	35223	Endocardial/interventricular septal pacing lead	CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A05	Mechanical Problem		
									A070909	Oversensing		
									A070910	Undersensing		
									A072202	Low Impedance		
									A090801	Signal Artifact/Noise		
							F12	Serious Injury/ Illness/ Impairment	A05	Mechanical Problem		
									A070909	Oversensing		
									A070910	Undersensing		
									A072202	Low Impedance		
									A090801	Signal Artifact/Noise		
							F23	Unexpected Medical Intervention	A05	Mechanical Problem		
									A070909	Oversensing		
									A070910	Undersensing		
									A072202	Low Impedance		
									A090801	Signal Artifact/Noise		

6971	2022	All other reportable incidents	17882	Physiologic-monitoring defibrillation system	CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A0710	Failure to Conduct		
							F26	No Health Consequences or Impact	A0710	Failure to Conduct		
6972	2022	Unanticipated serious deterioration in state of health	47106	Drug-eluting endocardial pacing lead	CT1684	Cardiac pacemakers and associated devices	E060101	Asystole	A070909	Oversensing	04/05/2022	
							F08	Hospitalization or Prolonged Hospitalization	A070909	Oversensing	04/05/2022	
							F19	Surgical Intervention	A070909	Oversensing	04/05/2022	
6973	2022	All other reportable incidents	47265	Dual-chamber implantable pacemaker, rate-responsive	CT141	Electrodes/Leads and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
					CT145	Cardiovascular devices	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
					CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
					CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
					CT233	Surgical	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
					CT2406	Implantable	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
					CT2407	Active implantable	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
					CT2727	Active implant devices/ancillaries	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
					CT2750	Implantable cardiac electrical stimulators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
					CT2985	Active	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
					CT3001	Implantable metal	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
					CT3002	Implantable synthetic polymer	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
					CT3006	Implantable inorganic material	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
					CT314	Electrically powered	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
					CT321	Long-term surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
					CT449	Generators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
					CT981	Single use	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
					CT983	Surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
6974	2022	All other reportable incidents	17882	Physiologic-monitoring defibrillation system	CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A07	Electrical /Electronic Property Problem		
							F27	Problem Identified During Non-Clinical Procedure	A07	Electrical /Electronic Property Problem		
6975	2022	All other reportable incidents	17882	Physiologic-monitoring defibrillation system	CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A07	Electrical /Electronic Property Problem		

							F27	Problem Identified During Non-Clinical Procedure	A07	Electrical /Electronic Property Problem		
6976	2022	All other reportable incidents	17882	Physiologic-monitoring defibrillation system	CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A07	Electrical /Electronic Property Problem		
							F27	Problem Identified During Non-Clinical Procedure	A07	Electrical /Electronic Property Problem		
6977	2022	All other reportable incidents	17882	Physiologic-monitoring defibrillation system	CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A07	Electrical /Electronic Property Problem		
							F27	Problem Identified During Non-Clinical Procedure	A07	Electrical /Electronic Property Problem		
6978	2022	All other reportable incidents	17882	Physiologic-monitoring defibrillation system	CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A07	Electrical /Electronic Property Problem		
							F27	Problem Identified During Non-Clinical Procedure	A07	Electrical /Electronic Property Problem		
6979	2022	All other reportable incidents	35852	Single-chamber implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	02/08/2022	02/08/2022
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	02/08/2022	02/08/2022
							F1903	Device Explantation	A070504	Premature Discharge of Battery	02/08/2022	02/08/2022
6980	2022	All other reportable incidents	17882	Physiologic-monitoring defibrillation system	CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A07	Electrical /Electronic Property Problem		
							F27	Problem Identified During Non-Clinical Procedure	A07	Electrical /Electronic Property Problem		
6981	2022	All other reportable incidents	17882	Physiologic-monitoring defibrillation system	CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A07	Electrical /Electronic Property Problem		
							F27	Problem Identified During Non-Clinical Procedure	A07	Electrical /Electronic Property Problem		
6982	2022	All other reportable incidents	35223	Endocardial/interventricular septal pacing lead	CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070909	Oversensing		
									A072202	Low Impedance		
							F1901	Additional Surgery	A070909	Oversensing		
									A072202	Low Impedance		
6983	2021	All other reportable incidents	17882	Physiologic-monitoring defibrillation system	CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A0719	Unexpected Shutdown		
							F27	Problem Identified During Non-Clinical Procedure	A0719	Unexpected Shutdown		
6984	2021	All other reportable incidents	36241	Pacemaker/defibrillator lead	CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A040101	Fracture	19/01/2022	19/01/2022
									A090801	Signal Artifact/Noise	19/01/2022	19/01/2022
							F08	Hospitalization or Prolonged Hospitalization	A040101	Fracture	19/01/2022	19/01/2022
									A090801	Signal Artifact/Noise	19/01/2022	19/01/2022
							F1905	Device Revision or Replacement	A040101	Fracture	19/01/2022	19/01/2022
									A090801	Signal Artifact/Noise	19/01/2022	19/01/2022
							F2203	Imaging Required	A040101	Fracture	19/01/2022	19/01/2022
									A090801	Signal Artifact/Noise	19/01/2022	19/01/2022
6985	2021	All other reportable incidents	47265	Dual-chamber implantable pacemaker, rate-responsive	CT141	Electrodes/Leads and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem	24/12/2021	24/12/2021
									A070504	Premature Discharge of Battery	24/12/2021	24/12/2021
							F08	Hospitalization or Prolonged Hospitalization	A0705	Battery Problem	24/12/2021	24/12/2021
									A070504	Premature Discharge of Battery	24/12/2021	24/12/2021
							F1905	Device Revision or Replacement	A0705	Battery Problem	24/12/2021	24/12/2021
									A070504	Premature Discharge of Battery	24/12/2021	24/12/2021
					CT145	Cardiovascular devices	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem	24/12/2021	24/12/2021
									A070504	Premature Discharge of Battery	24/12/2021	24/12/2021
							F08	Hospitalization or Prolonged Hospitalization	A0705	Battery Problem	24/12/2021	24/12/2021
									A070504	Premature Discharge of Battery	24/12/2021	24/12/2021
							F1905	Device Revision or Replacement	A0705	Battery Problem	24/12/2021	24/12/2021
									A070504	Premature Discharge of Battery	24/12/2021	24/12/2021
					CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem	24/12/2021	24/12/2021
									A070504	Premature Discharge of Battery	24/12/2021	24/12/2021
							F08	Hospitalization or Prolonged Hospitalization	A0705	Battery Problem	24/12/2021	24/12/2021
									A070504	Premature Discharge of Battery	24/12/2021	24/12/2021
							F1905	Device Revision or Replacement	A0705	Battery Problem	24/12/2021	24/12/2021
									A070504	Premature Discharge of Battery	24/12/2021	24/12/2021
					CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem	24/12/2021	24/12/2021
									A070504	Premature Discharge of Battery	24/12/2021	24/12/2021
							F08	Hospitalization or Prolonged Hospitalization	A0705	Battery Problem	24/12/2021	24/12/2021
									A070504	Premature Discharge of Battery	24/12/2021	24/12/2021
							F1905	Device Revision or Replacement	A0705	Battery Problem	24/12/2021	24/12/2021
									A070504	Premature Discharge of Battery	24/12/2021	24/12/2021
					CT233	Surgical	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem	24/12/2021	24/12/2021
									A070504	Premature Discharge of Battery	24/12/2021	24/12/2021
							F08	Hospitalization or Prolonged Hospitalization	A0705	Battery Problem	24/12/2021	24/12/2021
									A070504	Premature Discharge of Battery	24/12/2021	24/12/2021
							F1905	Device Revision or Replacement	A0705	Battery Problem	24/12/2021	24/12/2021
									A070504	Premature Discharge of Battery	24/12/2021	24/12/2021
					CT2406	Implantable	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem	24/12/2021	24/12/2021
									A070504	Premature Discharge of Battery	24/12/2021	24/12/2021
							F08	Hospitalization or Prolonged Hospitalization	A0705	Battery Problem	24/12/2021	24/12/2021
									A070504	Premature Discharge of Battery	24/12/2021	24/12/2021
							F1905	Device Revision or Replacement	A0705	Battery Problem	24/12/2021	24/12/2021
									A070504	Premature Discharge of Battery	24/12/2021	24/12/2021
					CT2407	Active implantable	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem	24/12/2021	24/12/2021
									A070504	Premature Discharge of Battery	24/12/2021	24/12/2021
							F08	Hospitalization or Prolonged Hospitalization	A0705	Battery Problem	24/12/2021	24/12/2021
									A070504	Premature Discharge of Battery	24/12/2021	24/12/2021
							F1905	Device Revision or Replacement	A0705	Battery Problem	24/12/2021	24/12/2021
									A070504	Premature Discharge of Battery	24/12/2021	24/12/2021
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem	24/12/2021	24/12/2021
									A070504	Premature Discharge of Battery	24/12/2021	24/12/2021
							F08	Hospitalization or Prolonged Hospitalization	A0705	Battery Problem	24/12/2021	24/12/2021
									A070504	Premature Discharge of Battery	24/12/2021	24/12/2021
							F1905	Device Revision or Replacement	A0705	Battery Problem	24/12/2021	24/12/2021
									A070504	Premature Discharge of Battery	24/12/2021	24/12/2021
					CT2727	Active implant devices/ancillaries	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem	24/12/2021	24/12/2021
									A070504	Premature Discharge of Battery	24/12/2021	24/12/2021
							F08	Hospitalization or Prolonged Hospitalization	A0705	Battery Problem	24/12/2021	24/12/2021
									A070504	Premature Discharge of Battery	24/12/2021	24/12/2021
							F1905	Device Revision or Replacement	A0705	Battery Problem	24/12/2021	24/12/2021
									A070504	Premature Discharge of Battery	24/12/2021	24/12/2021
					CT2750	Implantable cardiac electrical stimulators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem	24/12/2021	24/12/2021
									A070504	Premature Discharge of Battery	24/12/2021	24/12/2021
							F08	Hospitalization or Prolonged Hospitalization	A0705	Battery Problem	24/12/2021	24/12/2021
									A070504	Premature Discharge of Battery	24/12/2021	24/12/2021
							F1905	Device Revision or Replacement	A0705	Battery Problem	24/12/2021	24/12/2021
									A070504	Premature Discharge of Battery	24/12/2021	24/12/2021

					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem	24/12/2021	24/12/2021
									A070504	Premature Discharge of Battery	24/12/2021	24/12/2021
							F08	Hospitalization or Prolonged Hospitalization	A0705	Battery Problem	24/12/2021	24/12/2021
									A070504	Premature Discharge of Battery	24/12/2021	24/12/2021
							F1905	Device Revision or Replacement	A0705	Battery Problem	24/12/2021	24/12/2021
									A070504	Premature Discharge of Battery	24/12/2021	24/12/2021
					CT2985	Active	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem	24/12/2021	24/12/2021
									A070504	Premature Discharge of Battery	24/12/2021	24/12/2021
							F08	Hospitalization or Prolonged Hospitalization	A0705	Battery Problem	24/12/2021	24/12/2021
									A070504	Premature Discharge of Battery	24/12/2021	24/12/2021
							F1905	Device Revision or Replacement	A0705	Battery Problem	24/12/2021	24/12/2021
									A070504	Premature Discharge of Battery	24/12/2021	24/12/2021
					CT3001	Implantable metal	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem	24/12/2021	24/12/2021
									A070504	Premature Discharge of Battery	24/12/2021	24/12/2021
							F08	Hospitalization or Prolonged Hospitalization	A0705	Battery Problem	24/12/2021	24/12/2021
									A070504	Premature Discharge of Battery	24/12/2021	24/12/2021
							F1905	Device Revision or Replacement	A0705	Battery Problem	24/12/2021	24/12/2021
									A070504	Premature Discharge of Battery	24/12/2021	24/12/2021
					CT3002	Implantable synthetic polymer	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem	24/12/2021	24/12/2021
									A070504	Premature Discharge of Battery	24/12/2021	24/12/2021
							F08	Hospitalization or Prolonged Hospitalization	A0705	Battery Problem	24/12/2021	24/12/2021
									A070504	Premature Discharge of Battery	24/12/2021	24/12/2021
							F1905	Device Revision or Replacement	A0705	Battery Problem	24/12/2021	24/12/2021
									A070504	Premature Discharge of Battery	24/12/2021	24/12/2021
					CT3006	Implantable inorganic material	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem	24/12/2021	24/12/2021
									A070504	Premature Discharge of Battery	24/12/2021	24/12/2021
							F08	Hospitalization or Prolonged Hospitalization	A0705	Battery Problem	24/12/2021	24/12/2021
									A070504	Premature Discharge of Battery	24/12/2021	24/12/2021
							F1905	Device Revision or Replacement	A0705	Battery Problem	24/12/2021	24/12/2021
									A070504	Premature Discharge of Battery	24/12/2021	24/12/2021
					CT314	Electrically powered	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem	24/12/2021	24/12/2021
									A070504	Premature Discharge of Battery	24/12/2021	24/12/2021
							F08	Hospitalization or Prolonged Hospitalization	A0705	Battery Problem	24/12/2021	24/12/2021
									A070504	Premature Discharge of Battery	24/12/2021	24/12/2021
							F1905	Device Revision or Replacement	A0705	Battery Problem	24/12/2021	24/12/2021
									A070504	Premature Discharge of Battery	24/12/2021	24/12/2021
					CT321	Long-term surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem	24/12/2021	24/12/2021
									A070504	Premature Discharge of Battery	24/12/2021	24/12/2021
							F08	Hospitalization or Prolonged Hospitalization	A0705	Battery Problem	24/12/2021	24/12/2021
									A070504	Premature Discharge of Battery	24/12/2021	24/12/2021
							F1905	Device Revision or Replacement	A0705	Battery Problem	24/12/2021	24/12/2021
									A070504	Premature Discharge of Battery	24/12/2021	24/12/2021
					CT449	Generators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem	24/12/2021	24/12/2021
									A070504	Premature Discharge of Battery	24/12/2021	24/12/2021
							F08	Hospitalization or Prolonged Hospitalization	A0705	Battery Problem	24/12/2021	24/12/2021
									A070504	Premature Discharge of Battery	24/12/2021	24/12/2021
							F1905	Device Revision or Replacement	A0705	Battery Problem	24/12/2021	24/12/2021
									A070504	Premature Discharge of Battery	24/12/2021	24/12/2021
					CT981	Single use	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem	24/12/2021	24/12/2021
									A070504	Premature Discharge of Battery	24/12/2021	24/12/2021
							F08	Hospitalization or Prolonged Hospitalization	A0705	Battery Problem	24/12/2021	24/12/2021
									A070504	Premature Discharge of Battery	24/12/2021	24/12/2021
							F1905	Device Revision or Replacement	A0705	Battery Problem	24/12/2021	24/12/2021
									A070504	Premature Discharge of Battery	24/12/2021	24/12/2021
					CT983	Surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem	24/12/2021	24/12/2021
									A070504	Premature Discharge of Battery	24/12/2021	24/12/2021
							F08	Hospitalization or Prolonged Hospitalization	A0705	Battery Problem	24/12/2021	24/12/2021
									A070504	Premature Discharge of Battery	24/12/2021	24/12/2021
							F1905	Device Revision or Replacement	A0705	Battery Problem	24/12/2021	24/12/2021
									A070504	Premature Discharge of Battery	24/12/2021	24/12/2021
6986	2021	All other reportable incidents	37265	Dual-chamber implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A072201	High Impedance		
							F26	No Health Consequences or Impact	A072201	High Impedance		
6987	2021	All other reportable incidents	36241	Pacemaker/defibrillator lead	CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A072201	High Impedance		
							F26	No Health Consequences or Impact	A072201	High Impedance		
6988	2022	All other reportable incidents	36241	Pacemaker/defibrillator lead	CT1684	Cardiac pacemakers and associated devices	E0601	Arrhythmia	A070909	Oversensing		
									A0712	Pacing Problem		
									A090801	Signal Artifact/Noise		
							E2403	No Clinical Signs, Symptoms or Conditions	A070909	Oversensing		
									A0712	Pacing Problem		
									A090801	Signal Artifact/Noise		
							F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A070909	Oversensing		
									A0712	Pacing Problem		
									A090801	Signal Artifact/Noise		
6989	2022	All other reportable incidents	36241	Pacemaker/defibrillator lead	CT1684	Cardiac pacemakers and associated devices	E2104	Electric Shock	A071302	Inappropriate/Inadequate Shock/Stimulation	14/06/2022	14/06/2022
									A0908	Incorrect, Inadequate or Imprecise Result or Readings	14/06/2022	14/06/2022
							E2330	Pain	A071302	Inappropriate/Inadequate Shock/Stimulation	14/06/2022	14/06/2022
									A0908	Incorrect, Inadequate or Imprecise Result or Readings	14/06/2022	14/06/2022
							E2403	No Clinical Signs, Symptoms or Conditions	A071302	Inappropriate/Inadequate Shock/Stimulation	14/06/2022	14/06/2022
									A0908	Incorrect, Inadequate or Imprecise Result or Readings	14/06/2022	14/06/2022
							F08	Hospitalization or Prolonged Hospitalization	A071302	Inappropriate/Inadequate Shock/Stimulation	14/06/2022	14/06/2022
									A0908	Incorrect, Inadequate or Imprecise Result or Readings	14/06/2022	14/06/2022
							F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A071302	Inappropriate/Inadequate Shock/Stimulation	14/06/2022	14/06/2022
									A0908	Incorrect, Inadequate or Imprecise Result or Readings	14/06/2022	14/06/2022
							F1905	Device Revision or Replacement	A071302	Inappropriate/Inadequate Shock/Stimulation	14/06/2022	14/06/2022

									A0908	Incorrect, Inadequate or Imprecise Result or Readings	14/06/2022	14/06/2022
6990	2022	All other reportable incidents	37265	Dual-chamber implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
									A160601	Premature Elective Replacement Indicator		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
									A160601	Premature Elective Replacement Indicator		
6991	2021	All other reportable incidents	37265	Dual-chamber implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A070102	High Capture Threshold		
									A070909	Oversensing		
							F26	No Health Consequences or Impact	A070102	High Capture Threshold		
									A070909	Oversensing		
6992	2021	All other reportable incidents	47265	Dual-chamber implantable pacemaker, rate-responsive	CT141	Electrodes/Leads and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	25/11/2021	25/11/2021
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	25/11/2021	25/11/2021
							F1903	Device Explantation	A070504	Premature Discharge of Battery	25/11/2021	25/11/2021
					CT145	Cardiovascular devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	25/11/2021	25/11/2021
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	25/11/2021	25/11/2021
							F1903	Device Explantation	A070504	Premature Discharge of Battery	25/11/2021	25/11/2021
					CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	25/11/2021	25/11/2021
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	25/11/2021	25/11/2021
							F1903	Device Explantation	A070504	Premature Discharge of Battery	25/11/2021	25/11/2021
					CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	25/11/2021	25/11/2021
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	25/11/2021	25/11/2021
							F1903	Device Explantation	A070504	Premature Discharge of Battery	25/11/2021	25/11/2021
					CT233	Surgical	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	25/11/2021	25/11/2021
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	25/11/2021	25/11/2021
							F1903	Device Explantation	A070504	Premature Discharge of Battery	25/11/2021	25/11/2021
					CT2406	Implantable	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	25/11/2021	25/11/2021
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	25/11/2021	25/11/2021
							F1903	Device Explantation	A070504	Premature Discharge of Battery	25/11/2021	25/11/2021
					CT2407	Active implantable	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	25/11/2021	25/11/2021
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	25/11/2021	25/11/2021
							F1903	Device Explantation	A070504	Premature Discharge of Battery	25/11/2021	25/11/2021
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	25/11/2021	25/11/2021
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	25/11/2021	25/11/2021
							F1903	Device Explantation	A070504	Premature Discharge of Battery	25/11/2021	25/11/2021
					CT2727	Active implant devices/ancillaries	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	25/11/2021	25/11/2021
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	25/11/2021	25/11/2021
							F1903	Device Explantation	A070504	Premature Discharge of Battery	25/11/2021	25/11/2021
					CT2750	Implantable cardiac electrical stimulators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	25/11/2021	25/11/2021
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	25/11/2021	25/11/2021
							F1903	Device Explantation	A070504	Premature Discharge of Battery	25/11/2021	25/11/2021
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	25/11/2021	25/11/2021
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	25/11/2021	25/11/2021
							F1903	Device Explantation	A070504	Premature Discharge of Battery	25/11/2021	25/11/2021
					CT2985	Active	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	25/11/2021	25/11/2021
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	25/11/2021	25/11/2021
							F1903	Device Explantation	A070504	Premature Discharge of Battery	25/11/2021	25/11/2021
					CT3001	Implantable metal	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	25/11/2021	25/11/2021
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	25/11/2021	25/11/2021
							F1903	Device Explantation	A070504	Premature Discharge of Battery	25/11/2021	25/11/2021
					CT3002	Implantable synthetic polymer	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	25/11/2021	25/11/2021
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	25/11/2021	25/11/2021
							F1903	Device Explantation	A070504	Premature Discharge of Battery	25/11/2021	25/11/2021
					CT3006	Implantable inorganic material	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	25/11/2021	25/11/2021
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	25/11/2021	25/11/2021
							F1903	Device Explantation	A070504	Premature Discharge of Battery	25/11/2021	25/11/2021
					CT314	Electrically powered	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	25/11/2021	25/11/2021
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	25/11/2021	25/11/2021
							F1903	Device Explantation	A070504	Premature Discharge of Battery	25/11/2021	25/11/2021
					CT321	Long-term surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	25/11/2021	25/11/2021
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	25/11/2021	25/11/2021
							F1903	Device Explantation	A070504	Premature Discharge of Battery	25/11/2021	25/11/2021
					CT449	Generators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	25/11/2021	25/11/2021
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	25/11/2021	25/11/2021
							F1903	Device Explantation	A070504	Premature Discharge of Battery	25/11/2021	25/11/2021
					CT981	Single use	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	25/11/2021	25/11/2021
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	25/11/2021	25/11/2021
							F1903	Device Explantation	A070504	Premature Discharge of Battery	25/11/2021	25/11/2021
					CT983	Surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	25/11/2021	25/11/2021
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	25/11/2021	25/11/2021
							F1903	Device Explantation	A070504	Premature Discharge of Battery	25/11/2021	25/11/2021
6993	2021	All other reportable incidents	47267	Single-chamber implantable pacemaker, rate-responsive	CT141	Electrodes/Leads and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	16/06/2022	16/06/2022
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	16/06/2022	16/06/2022
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	16/06/2022	16/06/2022
					CT145	Cardiovascular devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	16/06/2022	16/06/2022
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	16/06/2022	16/06/2022
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	16/06/2022	16/06/2022
					CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	16/06/2022	16/06/2022
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	16/06/2022	16/06/2022
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	16/06/2022	16/06/2022
					CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	16/06/2022	16/06/2022
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	16/06/2022	16/06/2022
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	16/06/2022	16/06/2022
					CT233	Surgical	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	16/06/2022	16/06/2022
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	16/06/2022	16/06/2022
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	16/06/2022	16/06/2022
					CT2406	Implantable	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	16/06/2022	16/06/2022
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	16/06/2022	16/06/2022
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	16/06/2022	16/06/2022

					CT2407	Active implantable	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	16/06/2022	16/06/2022
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	16/06/2022	16/06/2022
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	16/06/2022	16/06/2022
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	16/06/2022	16/06/2022
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	16/06/2022	16/06/2022
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	16/06/2022	16/06/2022
					CT2727	Active implant devices/ancillaries	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	16/06/2022	16/06/2022
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	16/06/2022	16/06/2022
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	16/06/2022	16/06/2022
					CT2750	Implantable cardiac electrical stimulators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	16/06/2022	16/06/2022
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	16/06/2022	16/06/2022
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	16/06/2022	16/06/2022
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	16/06/2022	16/06/2022
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	16/06/2022	16/06/2022
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	16/06/2022	16/06/2022
					CT2985	Active	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	16/06/2022	16/06/2022
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	16/06/2022	16/06/2022
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	16/06/2022	16/06/2022
					CT3001	Implantable metal	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	16/06/2022	16/06/2022
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	16/06/2022	16/06/2022
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	16/06/2022	16/06/2022
					CT3002	Implantable synthetic polymer	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	16/06/2022	16/06/2022
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	16/06/2022	16/06/2022
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	16/06/2022	16/06/2022
					CT3006	Implantable inorganic material	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	16/06/2022	16/06/2022
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	16/06/2022	16/06/2022
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	16/06/2022	16/06/2022
					CT314	Electrically powered	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	16/06/2022	16/06/2022
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	16/06/2022	16/06/2022
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	16/06/2022	16/06/2022
					CT321	Long-term surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	16/06/2022	16/06/2022
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	16/06/2022	16/06/2022
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	16/06/2022	16/06/2022
					CT449	Generators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	16/06/2022	16/06/2022
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	16/06/2022	16/06/2022
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	16/06/2022	16/06/2022
					CT981	Single use	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	16/06/2022	16/06/2022
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	16/06/2022	16/06/2022
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	16/06/2022	16/06/2022
					CT983	Surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	16/06/2022	16/06/2022
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	16/06/2022	16/06/2022
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	16/06/2022	16/06/2022
6994	2021	All other reportable incidents	36241	Pacemaker/defibrillator lead	CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A072201	High Impedance		
							F26	No Health Consequences or Impact	A072201	High Impedance		
6995	2021	All other reportable incidents	36241	Pacemaker/defibrillator lead	CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A072202	Low Impedance		
							F11	Non-Serious Injury/ Illness/ Impairment	A072202	Low Impedance		
6996	2021	All other reportable incidents	36241	Pacemaker/defibrillator lead	CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070101	Failure to Capture		
									A070102	High Capture Threshold		
									A070909	Oversensing		
									A072201	High Impedance		
									A090801	Signal Artifact/Noise		
							F23	Unexpected Medical Intervention	A070101	Failure to Capture		
									A070102	High Capture Threshold		
									A070909	Oversensing		
									A072201	High Impedance		
									A090801	Signal Artifact/Noise		
6997	2022	All other reportable incidents	35852	Single-chamber implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem	04/08/2022	04/08/2022
									A070504	Premature Discharge of Battery	04/08/2022	04/08/2022
							F08	Hospitalization or Prolonged Hospitalization	A0705	Battery Problem	04/08/2022	04/08/2022
									A070504	Premature Discharge of Battery	04/08/2022	04/08/2022
							F1905	Device Revision or Replacement	A0705	Battery Problem	04/08/2022	04/08/2022
									A070504	Premature Discharge of Battery	04/08/2022	04/08/2022
6998	2021	All other reportable incidents	37265	Dual-chamber implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem	24/11/2021	24/11/2021
									A070504	Premature Discharge of Battery	24/11/2021	24/11/2021
							F08	Hospitalization or Prolonged Hospitalization	A0705	Battery Problem	24/11/2021	24/11/2021
									A070504	Premature Discharge of Battery	24/11/2021	24/11/2021
							F1905	Device Revision or Replacement	A0705	Battery Problem	24/11/2021	24/11/2021
									A070504	Premature Discharge of Battery	24/11/2021	24/11/2021
6999	2021	All other reportable incidents	47267	Single-chamber implantable pacemaker, rate-responsive	CT141	Electrodes/Leads and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
					CT145	Cardiovascular devices	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
					CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
					CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
					CT233	Surgical	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		

					CT2406	Implantable	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
					CT2407	Active implantable	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
					CT2727	Active implant devices/ancillaries	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
					CT2750	Implantable cardiac electrical stimulators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
					CT2985	Active	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
					CT3001	Implantable metal	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
					CT3002	Implantable synthetic polymer	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
					CT3006	Implantable inorganic material	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
					CT314	Electrically powered	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
					CT321	Long-term surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
					CT449	Generators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
					CT981	Single use	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
					CT983	Surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
7000	2021	All other reportable incidents	35223	Endocardial/interventricular septal pacing lead	CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070102	High Capture Threshold		
									A090801	Signal Artifact/Noise		
							F26	No Health Consequences or Impact	A070102	High Capture Threshold		
									A090801	Signal Artifact/Noise		
7001	2021	All other reportable incidents	35039	Epicardial pacing lead	CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A040101	Fracture		
									A070101	Failure to Capture		
									A070102	High Capture Threshold		
									A070103	Intermittent Capture		
									A0722	Impedance Problem		
									A072201	High Impedance		
							F26	No Health Consequences or Impact	A040101	Fracture		
									A070101	Failure to Capture		
									A070102	High Capture Threshold		
									A070103	Intermittent Capture		
									A0722	Impedance Problem		
									A072201	High Impedance		
7002	2021	All other reportable incidents	35223	Endocardial/interventricular septal pacing lead	CT1684	Cardiac pacemakers and associated devices	E011903	Syncope/Fainting	A070101	Failure to Capture	29/10/2021	29/10/2021
									A070104	Unstable Capture Threshold	29/10/2021	29/10/2021
									A070901	Decreased Sensitivity	29/10/2021	29/10/2021
							E2328	Obstruction/Occlusion	A070101	Failure to Capture	29/10/2021	29/10/2021
									A070104	Unstable Capture Threshold	29/10/2021	29/10/2021
									A070901	Decreased Sensitivity	29/10/2021	29/10/2021
							E2403	No Clinical Signs, Symptoms or Conditions	A070101	Failure to Capture	29/10/2021	29/10/2021
									A070104	Unstable Capture Threshold	29/10/2021	29/10/2021
									A070901	Decreased Sensitivity	29/10/2021	29/10/2021
							F1203	Life Threatening Illness or Injury	A070101	Failure to Capture	29/10/2021	29/10/2021
									A070104	Unstable Capture Threshold	29/10/2021	29/10/2021
									A070901	Decreased Sensitivity	29/10/2021	29/10/2021

							F23	Unexpected Medical Intervention	A070101	Failure to Capture	29/10/2021	29/10/2021
									A070104	Unstable Capture Threshold	29/10/2021	29/10/2021
									A070901	Decreased Sensitivity	29/10/2021	29/10/2021
7003	2022	All other reportable incidents	47265	Dual-chamber implantable pacemaker, rate-responsive	CT141	Electrodes/Leads and associated devices	E0112	Dizziness	A070504	Premature Discharge of Battery	21/06/2022	21/06/2022
									A071102	Failure to Interrogate	21/06/2022	21/06/2022
							E0601	Arrhythmia	A070504	Premature Discharge of Battery	21/06/2022	21/06/2022
									A071102	Failure to Interrogate	21/06/2022	21/06/2022
							E0717	Dyspnea	A070504	Premature Discharge of Battery	21/06/2022	21/06/2022
									A071102	Failure to Interrogate	21/06/2022	21/06/2022
							E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	21/06/2022	21/06/2022
									A071102	Failure to Interrogate	21/06/2022	21/06/2022
							F1901	Additional Surgery	A070504	Premature Discharge of Battery	21/06/2022	21/06/2022
									A071102	Failure to Interrogate	21/06/2022	21/06/2022
					CT145	Cardiovascular devices	E0112	Dizziness	A070504	Premature Discharge of Battery	21/06/2022	21/06/2022
									A071102	Failure to Interrogate	21/06/2022	21/06/2022
							E0601	Arrhythmia	A070504	Premature Discharge of Battery	21/06/2022	21/06/2022
									A071102	Failure to Interrogate	21/06/2022	21/06/2022
							E0717	Dyspnea	A070504	Premature Discharge of Battery	21/06/2022	21/06/2022
									A071102	Failure to Interrogate	21/06/2022	21/06/2022
							E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	21/06/2022	21/06/2022
									A071102	Failure to Interrogate	21/06/2022	21/06/2022
							F1901	Additional Surgery	A070504	Premature Discharge of Battery	21/06/2022	21/06/2022
									A071102	Failure to Interrogate	21/06/2022	21/06/2022
					CT1684	Cardiac pacemakers and associated devices	E0112	Dizziness	A070504	Premature Discharge of Battery	21/06/2022	21/06/2022
									A071102	Failure to Interrogate	21/06/2022	21/06/2022
							E0601	Arrhythmia	A070504	Premature Discharge of Battery	21/06/2022	21/06/2022
									A071102	Failure to Interrogate	21/06/2022	21/06/2022
							E0717	Dyspnea	A070504	Premature Discharge of Battery	21/06/2022	21/06/2022
									A071102	Failure to Interrogate	21/06/2022	21/06/2022
							E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	21/06/2022	21/06/2022
									A071102	Failure to Interrogate	21/06/2022	21/06/2022
							F1901	Additional Surgery	A070504	Premature Discharge of Battery	21/06/2022	21/06/2022
									A071102	Failure to Interrogate	21/06/2022	21/06/2022
					CT1685	Cardiac pacemakers	E0112	Dizziness	A070504	Premature Discharge of Battery	21/06/2022	21/06/2022
									A071102	Failure to Interrogate	21/06/2022	21/06/2022
							E0601	Arrhythmia	A070504	Premature Discharge of Battery	21/06/2022	21/06/2022
									A071102	Failure to Interrogate	21/06/2022	21/06/2022
							E0717	Dyspnea	A070504	Premature Discharge of Battery	21/06/2022	21/06/2022
									A071102	Failure to Interrogate	21/06/2022	21/06/2022
							E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	21/06/2022	21/06/2022
									A071102	Failure to Interrogate	21/06/2022	21/06/2022
							F1901	Additional Surgery	A070504	Premature Discharge of Battery	21/06/2022	21/06/2022
									A071102	Failure to Interrogate	21/06/2022	21/06/2022
					CT233	Surgical	E0112	Dizziness	A070504	Premature Discharge of Battery	21/06/2022	21/06/2022
									A071102	Failure to Interrogate	21/06/2022	21/06/2022
							E0601	Arrhythmia	A070504	Premature Discharge of Battery	21/06/2022	21/06/2022
									A071102	Failure to Interrogate	21/06/2022	21/06/2022
							E0717	Dyspnea	A070504	Premature Discharge of Battery	21/06/2022	21/06/2022
									A071102	Failure to Interrogate	21/06/2022	21/06/2022
							E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	21/06/2022	21/06/2022
									A071102	Failure to Interrogate	21/06/2022	21/06/2022
							F1901	Additional Surgery	A070504	Premature Discharge of Battery	21/06/2022	21/06/2022
									A071102	Failure to Interrogate	21/06/2022	21/06/2022
					CT2406	Implantable	E0112	Dizziness	A070504	Premature Discharge of Battery	21/06/2022	21/06/2022
									A071102	Failure to Interrogate	21/06/2022	21/06/2022
							E0601	Arrhythmia	A070504	Premature Discharge of Battery	21/06/2022	21/06/2022
									A071102	Failure to Interrogate	21/06/2022	21/06/2022
							E0717	Dyspnea	A070504	Premature Discharge of Battery	21/06/2022	21/06/2022
									A071102	Failure to Interrogate	21/06/2022	21/06/2022
							E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	21/06/2022	21/06/2022
									A071102	Failure to Interrogate	21/06/2022	21/06/2022
							F1901	Additional Surgery	A070504	Premature Discharge of Battery	21/06/2022	21/06/2022
									A071102	Failure to Interrogate	21/06/2022	21/06/2022
					CT2407	Active implantable	E0112	Dizziness	A070504	Premature Discharge of Battery	21/06/2022	21/06/2022
									A071102	Failure to Interrogate	21/06/2022	21/06/2022
							E0601	Arrhythmia	A070504	Premature Discharge of Battery	21/06/2022	21/06/2022
									A071102	Failure to Interrogate	21/06/2022	21/06/2022
							E0717	Dyspnea	A070504	Premature Discharge of Battery	21/06/2022	21/06/2022
									A071102	Failure to Interrogate	21/06/2022	21/06/2022
							E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	21/06/2022	21/06/2022
									A071102	Failure to Interrogate	21/06/2022	21/06/2022
							F1901	Additional Surgery	A070504	Premature Discharge of Battery	21/06/2022	21/06/2022
									A071102	Failure to Interrogate	21/06/2022	21/06/2022
					CT257	Implantable cardiac pacemakers	E0112	Dizziness	A070504	Premature Discharge of Battery	21/06/2022	21/06/2022
									A071102	Failure to Interrogate	21/06/2022	21/06/2022
							E0601	Arrhythmia	A070504	Premature Discharge of Battery	21/06/2022	21/06/2022
									A071102	Failure to Interrogate	21/06/2022	21/06/2022
							E0717	Dyspnea	A070504	Premature Discharge of Battery	21/06/2022	21/06/2022
									A071102	Failure to Interrogate	21/06/2022	21/06/2022
							E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	21/06/2022	21/06/2022
									A071102	Failure to Interrogate	21/06/2022	21/06/2022
							F1901	Additional Surgery	A070504	Premature Discharge of Battery	21/06/2022	21/06/2022
									A071102	Failure to Interrogate	21/06/2022	21/06/2022
					CT2727	Active implant devices/ancillaries	E0112	Dizziness	A070504	Premature Discharge of Battery	21/06/2022	21/06/2022
									A071102	Failure to Interrogate	21/06/2022	21/06/2022
							E0601	Arrhythmia	A070504	Premature Discharge of Battery	21/06/2022	21/06/2022
									A071102	Failure to Interrogate	21/06/2022	21/06/2022
							E0717	Dyspnea	A070504	Premature Discharge of Battery	21/06/2022	21/06/2022

								A071102	Failure to Interrogate	21/06/2022	21/06/2022
						E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	21/06/2022	21/06/2022
								A071102	Failure to Interrogate	21/06/2022	21/06/2022
						F1901	Additional Surgery	A070504	Premature Discharge of Battery	21/06/2022	21/06/2022
								A071102	Failure to Interrogate	21/06/2022	21/06/2022
					CT2750	Implantable cardiac electrical stimulators and associated devices	E0112	Dizziness	A070504	Premature Discharge of Battery	21/06/2022	21/06/2022
									A071102	Failure to Interrogate	21/06/2022	21/06/2022
							E0601	Arrhythmia	A070504	Premature Discharge of Battery	21/06/2022	21/06/2022
									A071102	Failure to Interrogate	21/06/2022	21/06/2022
							E0717	Dyspnea	A070504	Premature Discharge of Battery	21/06/2022	21/06/2022
									A071102	Failure to Interrogate	21/06/2022	21/06/2022
							E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	21/06/2022	21/06/2022
									A071102	Failure to Interrogate	21/06/2022	21/06/2022
							F1901	Additional Surgery	A070504	Premature Discharge of Battery	21/06/2022	21/06/2022
									A071102	Failure to Interrogate	21/06/2022	21/06/2022
					CT2866	Implantable cardiac electrical stimulators	E0112	Dizziness	A070504	Premature Discharge of Battery	21/06/2022	21/06/2022
									A071102	Failure to Interrogate	21/06/2022	21/06/2022
							E0601	Arrhythmia	A070504	Premature Discharge of Battery	21/06/2022	21/06/2022
									A071102	Failure to Interrogate	21/06/2022	21/06/2022
							E0717	Dyspnea	A070504	Premature Discharge of Battery	21/06/2022	21/06/2022
									A071102	Failure to Interrogate	21/06/2022	21/06/2022
							E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	21/06/2022	21/06/2022
									A071102	Failure to Interrogate	21/06/2022	21/06/2022
							F1901	Additional Surgery	A070504	Premature Discharge of Battery	21/06/2022	21/06/2022
									A071102	Failure to Interrogate	21/06/2022	21/06/2022
					CT2985	Active	E0112	Dizziness	A070504	Premature Discharge of Battery	21/06/2022	21/06/2022
									A071102	Failure to Interrogate	21/06/2022	21/06/2022
							E0601	Arrhythmia	A070504	Premature Discharge of Battery	21/06/2022	21/06/2022
									A071102	Failure to Interrogate	21/06/2022	21/06/2022
							E0717	Dyspnea	A070504	Premature Discharge of Battery	21/06/2022	21/06/2022
									A071102	Failure to Interrogate	21/06/2022	21/06/2022
							E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	21/06/2022	21/06/2022
									A071102	Failure to Interrogate	21/06/2022	21/06/2022
							F1901	Additional Surgery	A070504	Premature Discharge of Battery	21/06/2022	21/06/2022
									A071102	Failure to Interrogate	21/06/2022	21/06/2022
					CT3001	Implantable metal	E0112	Dizziness	A070504	Premature Discharge of Battery	21/06/2022	21/06/2022
									A071102	Failure to Interrogate	21/06/2022	21/06/2022
							E0601	Arrhythmia	A070504	Premature Discharge of Battery	21/06/2022	21/06/2022
									A071102	Failure to Interrogate	21/06/2022	21/06/2022
							E0717	Dyspnea	A070504	Premature Discharge of Battery	21/06/2022	21/06/2022
									A071102	Failure to Interrogate	21/06/2022	21/06/2022
							E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	21/06/2022	21/06/2022
									A071102	Failure to Interrogate	21/06/2022	21/06/2022
							F1901	Additional Surgery	A070504	Premature Discharge of Battery	21/06/2022	21/06/2022
									A071102	Failure to Interrogate	21/06/2022	21/06/2022
					CT3002	Implantable synthetic polymer	E0112	Dizziness	A070504	Premature Discharge of Battery	21/06/2022	21/06/2022
									A071102	Failure to Interrogate	21/06/2022	21/06/2022
							E0601	Arrhythmia	A070504	Premature Discharge of Battery	21/06/2022	21/06/2022
									A071102	Failure to Interrogate	21/06/2022	21/06/2022
							E0717	Dyspnea	A070504	Premature Discharge of Battery	21/06/2022	21/06/2022
									A071102	Failure to Interrogate	21/06/2022	21/06/2022
							E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	21/06/2022	21/06/2022
									A071102	Failure to Interrogate	21/06/2022	21/06/2022
							F1901	Additional Surgery	A070504	Premature Discharge of Battery	21/06/2022	21/06/2022
									A071102	Failure to Interrogate	21/06/2022	21/06/2022
					CT3006	Implantable inorganic material	E0112	Dizziness	A070504	Premature Discharge of Battery	21/06/2022	21/06/2022
									A071102	Failure to Interrogate	21/06/2022	21/06/2022
							E0601	Arrhythmia	A070504	Premature Discharge of Battery	21/06/2022	21/06/2022
									A071102	Failure to Interrogate	21/06/2022	21/06/2022
							E0717	Dyspnea	A070504	Premature Discharge of Battery	21/06/2022	21/06/2022
									A071102	Failure to Interrogate	21/06/2022	21/06/2022
							E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	21/06/2022	21/06/2022
									A071102	Failure to Interrogate	21/06/2022	21/06/2022
							F1901	Additional Surgery	A070504	Premature Discharge of Battery	21/06/2022	21/06/2022
									A071102	Failure to Interrogate	21/06/2022	21/06/2022
					CT314	Electrically powered	E0112	Dizziness	A070504	Premature Discharge of Battery	21/06/2022	21/06/2022
									A071102	Failure to Interrogate	21/06/2022	21/06/2022
							E0601	Arrhythmia	A070504	Premature Discharge of Battery	21/06/2022	21/06/2022
									A071102	Failure to Interrogate	21/06/2022	21/06/2022
							E0717	Dyspnea	A070504	Premature Discharge of Battery	21/06/2022	21/06/2022
									A071102	Failure to Interrogate	21/06/2022	21/06/2022
							E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	21/06/2022	21/06/2022
									A071102	Failure to Interrogate	21/06/2022	21/06/2022
							F1901	Additional Surgery	A070504	Premature Discharge of Battery	21/06/2022	21/06/2022
									A071102	Failure to Interrogate	21/06/2022	21/06/2022
					CT321	Long-term surgical invasive	E0112	Dizziness	A070504	Premature Discharge of Battery	21/06/2022	21/06/2022
									A071102	Failure to Interrogate	21/06/2022	21/06/2022
							E0601	Arrhythmia	A070504	Premature Discharge of Battery	21/06/2022	21/06/2022
									A071102	Failure to Interrogate	21/06/2022	21/06/2022
							E0717	Dyspnea	A070504	Premature Discharge of Battery	21/06/2022	21/06/2022
									A071102	Failure to Interrogate	21/06/2022	21/06/2022
							E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	21/06/2022	21/06/2022
									A071102	Failure to Interrogate	21/06/2022	21/06/2022
							F1901	Additional Surgery	A070504	Premature Discharge of Battery	21/06/2022	21/06/2022
									A071102	Failure to Interrogate	21/06/2022	21/06/2022
					CT449	Generators and associated devices	E0112	Dizziness	A070504	Premature Discharge of Battery	21/06/2022	21/06/2022
									A071102	Failure to Interrogate	21/06/2022	21/06/2022
							E0601	Arrhythmia	A070504	Premature Discharge of Battery	21/06/2022	21/06/2022

								A071102	Failure to Interrogate	21/06/2022	21/06/2022
						E0717	Dyspnea	A070504	Premature Discharge of Battery	21/06/2022	21/06/2022
								A071102	Failure to Interrogate	21/06/2022	21/06/2022
						E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	21/06/2022	21/06/2022
								A071102	Failure to Interrogate	21/06/2022	21/06/2022
						F1901	Additional Surgery	A070504	Premature Discharge of Battery	21/06/2022	21/06/2022
								A071102	Failure to Interrogate	21/06/2022	21/06/2022
				CT981	Single use	E0112	Dizziness	A070504	Premature Discharge of Battery	21/06/2022	21/06/2022
								A071102	Failure to Interrogate	21/06/2022	21/06/2022
						E0601	Arrhythmia	A070504	Premature Discharge of Battery	21/06/2022	21/06/2022
								A071102	Failure to Interrogate	21/06/2022	21/06/2022
						E0717	Dyspnea	A070504	Premature Discharge of Battery	21/06/2022	21/06/2022
								A071102	Failure to Interrogate	21/06/2022	21/06/2022
						E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	21/06/2022	21/06/2022
								A071102	Failure to Interrogate	21/06/2022	21/06/2022
						F1901	Additional Surgery	A070504	Premature Discharge of Battery	21/06/2022	21/06/2022
								A071102	Failure to Interrogate	21/06/2022	21/06/2022
				CT983	Surgical invasive	E0112	Dizziness	A070504	Premature Discharge of Battery	21/06/2022	21/06/2022
								A071102	Failure to Interrogate	21/06/2022	21/06/2022
						E0601	Arrhythmia	A070504	Premature Discharge of Battery	21/06/2022	21/06/2022
								A071102	Failure to Interrogate	21/06/2022	21/06/2022
						E0717	Dyspnea	A070504	Premature Discharge of Battery	21/06/2022	21/06/2022
								A071102	Failure to Interrogate	21/06/2022	21/06/2022
						E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	21/06/2022	21/06/2022
								A071102	Failure to Interrogate	21/06/2022	21/06/2022
						F1901	Additional Surgery	A070504	Premature Discharge of Battery	21/06/2022	21/06/2022
								A071102	Failure to Interrogate	21/06/2022	21/06/2022
7004	2021	All other reportable incidents	47265	Dual-chamber implantable pacemaker, rate-responsive	CT141	Electrodes/Leads and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	11/11/2021	11/11/2021
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	11/11/2021	11/11/2021
							F1903	Device Explantation	A070504	Premature Discharge of Battery	11/11/2021	11/11/2021
					CT145	Cardiovascular devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	11/11/2021	11/11/2021
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	11/11/2021	11/11/2021
							F1903	Device Explantation	A070504	Premature Discharge of Battery	11/11/2021	11/11/2021
					CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	11/11/2021	11/11/2021
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	11/11/2021	11/11/2021
							F1903	Device Explantation	A070504	Premature Discharge of Battery	11/11/2021	11/11/2021
					CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	11/11/2021	11/11/2021
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	11/11/2021	11/11/2021
							F1903	Device Explantation	A070504	Premature Discharge of Battery	11/11/2021	11/11/2021
					CT233	Surgical	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	11/11/2021	11/11/2021
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	11/11/2021	11/11/2021
							F1903	Device Explantation	A070504	Premature Discharge of Battery	11/11/2021	11/11/2021
					CT2406	Implantable	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	11/11/2021	11/11/2021
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	11/11/2021	11/11/2021
							F1903	Device Explantation	A070504	Premature Discharge of Battery	11/11/2021	11/11/2021
					CT2407	Active implantable	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	11/11/2021	11/11/2021
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	11/11/2021	11/11/2021
							F1903	Device Explantation	A070504	Premature Discharge of Battery	11/11/2021	11/11/2021
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	11/11/2021	11/11/2021
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	11/11/2021	11/11/2021
							F1903	Device Explantation	A070504	Premature Discharge of Battery	11/11/2021	11/11/2021
					CT2727	Active implant devices/ancillaries	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	11/11/2021	11/11/2021
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	11/11/2021	11/11/2021
							F1903	Device Explantation	A070504	Premature Discharge of Battery	11/11/2021	11/11/2021
					CT2750	Implantable cardiac electrical stimulators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	11/11/2021	11/11/2021
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	11/11/2021	11/11/2021
							F1903	Device Explantation	A070504	Premature Discharge of Battery	11/11/2021	11/11/2021
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	11/11/2021	11/11/2021
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	11/11/2021	11/11/2021
							F1903	Device Explantation	A070504	Premature Discharge of Battery	11/11/2021	11/11/2021
					CT2985	Active	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	11/11/2021	11/11/2021
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	11/11/2021	11/11/2021
							F1903	Device Explantation	A070504	Premature Discharge of Battery	11/11/2021	11/11/2021
					CT3001	Implantable metal	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	11/11/2021	11/11/2021
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	11/11/2021	11/11/2021
							F1903	Device Explantation	A070504	Premature Discharge of Battery	11/11/2021	11/11/2021
					CT3002	Implantable synthetic polymer	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	11/11/2021	11/11/2021
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	11/11/2021	11/11/2021
							F1903	Device Explantation	A070504	Premature Discharge of Battery	11/11/2021	11/11/2021
					CT3006	Implantable inorganic material	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	11/11/2021	11/11/2021
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	11/11/2021	11/11/2021
							F1903	Device Explantation	A070504	Premature Discharge of Battery	11/11/2021	11/11/2021
					CT314	Electrically powered	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	11/11/2021	11/11/2021
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	11/11/2021	11/11/2021
							F1903	Device Explantation	A070504	Premature Discharge of Battery	11/11/2021	11/11/2021
					CT321	Long-term surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	11/11/2021	11/11/2021
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	11/11/2021	11/11/2021
							F1903	Device Explantation	A070504	Premature Discharge of Battery	11/11/2021	11/11/2021
					CT449	Generators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	11/11/2021	11/11/2021
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	11/11/2021	11/11/2021
							F1903	Device Explantation	A070504	Premature Discharge of Battery	11/11/2021	11/11/2021
					CT981	Single use	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	11/11/2021	11/11/2021
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	11/11/2021	11/11/2021
							F1903	Device Explantation	A070504	Premature Discharge of Battery	11/11/2021	11/11/2021
					CT983	Surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	11/11/2021	11/11/2021
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	11/11/2021	11/11/2021
							F1903	Device Explantation	A070504	Premature Discharge of Battery	11/11/2021	11/11/2021
7005	2021	All other reportable incidents	17882	Physiologic-monitoring defibrillation system	CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070803	Failure to Power Up		

									A0719	Unexpected Shutdown		
							F26	No Health Consequences or Impact	A070803	Failure to Power Up		
									A0719	Unexpected Shutdown		
7006	2022	All other reportable incidents	35223	Endocardial/interventricular septal pacing lead	CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A0401	Break		
									A040101	Fracture		
									A070909	Oversensing		
									A090801	Signal Artifact/Noise		
							F26	No Health Consequences or Impact	A0401	Break		
									A040101	Fracture		
									A070909	Oversensing		
									A090801	Signal Artifact/Noise		
7007	2021	All other reportable incidents	47267	Single-chamber implantable pacemaker, rate-responsive	CT141	Electrodes/Leads and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A09	Output Problem		
							F26	No Health Consequences or Impact	A09	Output Problem		
					CT145	Cardiovascular devices	E2403	No Clinical Signs, Symptoms or Conditions	A09	Output Problem		
							F26	No Health Consequences or Impact	A09	Output Problem		
					CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A09	Output Problem		
							F26	No Health Consequences or Impact	A09	Output Problem		
					CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A09	Output Problem		
							F26	No Health Consequences or Impact	A09	Output Problem		
					CT233	Surgical	E2403	No Clinical Signs, Symptoms or Conditions	A09	Output Problem		
							F26	No Health Consequences or Impact	A09	Output Problem		
					CT2406	Implantable	E2403	No Clinical Signs, Symptoms or Conditions	A09	Output Problem		
							F26	No Health Consequences or Impact	A09	Output Problem		
					CT2407	Active implantable	E2403	No Clinical Signs, Symptoms or Conditions	A09	Output Problem		
							F26	No Health Consequences or Impact	A09	Output Problem		
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A09	Output Problem		
							F26	No Health Consequences or Impact	A09	Output Problem		
					CT2727	Active implant devices/ancillaries	E2403	No Clinical Signs, Symptoms or Conditions	A09	Output Problem		
							F26	No Health Consequences or Impact	A09	Output Problem		
					CT2750	Implantable cardiac electrical stimulators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A09	Output Problem		
							F26	No Health Consequences or Impact	A09	Output Problem		
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A09	Output Problem		
							F26	No Health Consequences or Impact	A09	Output Problem		
					CT2985	Active	E2403	No Clinical Signs, Symptoms or Conditions	A09	Output Problem		
							F26	No Health Consequences or Impact	A09	Output Problem		
					CT3001	Implantable metal	E2403	No Clinical Signs, Symptoms or Conditions	A09	Output Problem		
							F26	No Health Consequences or Impact	A09	Output Problem		
					CT3002	Implantable synthetic polymer	E2403	No Clinical Signs, Symptoms or Conditions	A09	Output Problem		
							F26	No Health Consequences or Impact	A09	Output Problem		
					CT3006	Implantable inorganic material	E2403	No Clinical Signs, Symptoms or Conditions	A09	Output Problem		
							F26	No Health Consequences or Impact	A09	Output Problem		
					CT314	Electrically powered	E2403	No Clinical Signs, Symptoms or Conditions	A09	Output Problem		
							F26	No Health Consequences or Impact	A09	Output Problem		
					CT321	Long-term surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A09	Output Problem		
							F26	No Health Consequences or Impact	A09	Output Problem		
					CT449	Generators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A09	Output Problem		
							F26	No Health Consequences or Impact	A09	Output Problem		
					CT981	Single use	E2403	No Clinical Signs, Symptoms or Conditions	A09	Output Problem		
							F26	No Health Consequences or Impact	A09	Output Problem		
					CT983	Surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A09	Output Problem		
							F26	No Health Consequences or Impact	A09	Output Problem		
7008	2022	All other reportable incidents	35223	Endocardial/interventricular septal pacing lead	CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070901	Decreased Sensitivity		
									A070909	Oversensing		
							F26	No Health Consequences or Impact	A070901	Decreased Sensitivity		
									A070909	Oversensing		
7009	2021	All other reportable incidents	17882	Physiologic-monitoring defibrillation system	CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A0719	Unexpected Shutdown		
							F26	No Health Consequences or Impact	A0719	Unexpected Shutdown		
7010	2022	All other reportable incidents	17882	Physiologic-monitoring defibrillation system	CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A0719	Unexpected Shutdown		
							F27	Problem Identified During Non-Clinical Procedure	A0719	Unexpected Shutdown		
7011	2022	All other reportable incidents	17882	Physiologic-monitoring defibrillation system	CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A07	Electrical /Electronic Property Problem		
							F27	Problem Identified During Non-Clinical Procedure	A07	Electrical /Electronic Property Problem		
7012	2022	All other reportable incidents	17882	Physiologic-monitoring defibrillation system	CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A07	Electrical /Electronic Property Problem		
							F27	Problem Identified During Non-Clinical Procedure	A07	Electrical /Electronic Property Problem		
7013	2022	All other reportable incidents	17882	Physiologic-monitoring defibrillation system	CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A07	Electrical /Electronic Property Problem		
							F27	Problem Identified During Non-Clinical Procedure	A07	Electrical /Electronic Property Problem		
7014	2022	All other reportable incidents	47265	Dual-chamber implantable pacemaker, rate-responsive	CT141	Electrodes/Leads and associated devices	E0112	Dizziness	A070504	Premature Discharge of Battery		
									A071102	Failure to Interrogate		
							E0601	Arrhythmia	A070504	Premature Discharge of Battery		
									A071102	Failure to Interrogate		
							E0717	Dyspnea	A070504	Premature Discharge of Battery		
									A071102	Failure to Interrogate		
							E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
									A071102	Failure to Interrogate		
							F12	Serious Injury/ Illness/ Impairment	A070504	Premature Discharge of Battery		
									A071102	Failure to Interrogate		
					CT145	Cardiovascular devices	E0112	Dizziness	A070504	Premature Discharge of Battery		
									A071102	Failure to Interrogate		
							E0601	Arrhythmia	A070504	Premature Discharge of Battery		
									A071102	Failure to Interrogate		
							E0717	Dyspnea	A070504	Premature Discharge of Battery		
									A071102	Failure to Interrogate		
							E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
									A071102	Failure to Interrogate		
							F12	Serious Injury/ Illness/ Impairment	A070504	Premature Discharge of Battery		
									A071102	Failure to Interrogate		
					CT1684	Cardiac pacemakers and associated devices	E0112	Dizziness	A070504	Premature Discharge of Battery		
									A071102	Failure to Interrogate		
							E0601	Arrhythmia	A070504	Premature Discharge of Battery		

									A071102	Failure to Interrogate		
							E0717	Dyspnea	A070504	Premature Discharge of Battery		
									A071102	Failure to Interrogate		
							E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
									A071102	Failure to Interrogate		
							F12	Serious Injury/ Illness/ Impairment	A070504	Premature Discharge of Battery		
									A071102	Failure to Interrogate		
					CT1685	Cardiac pacemakers	E0112	Dizziness	A070504	Premature Discharge of Battery		
									A071102	Failure to Interrogate		
							E0601	Arrhythmia	A070504	Premature Discharge of Battery		
									A071102	Failure to Interrogate		
							E0717	Dyspnea	A070504	Premature Discharge of Battery		
									A071102	Failure to Interrogate		
							E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
									A071102	Failure to Interrogate		
							F12	Serious Injury/ Illness/ Impairment	A070504	Premature Discharge of Battery		
									A071102	Failure to Interrogate		
					CT233	Surgical	E0112	Dizziness	A070504	Premature Discharge of Battery		
									A071102	Failure to Interrogate		
							E0601	Arrhythmia	A070504	Premature Discharge of Battery		
									A071102	Failure to Interrogate		
							E0717	Dyspnea	A070504	Premature Discharge of Battery		
									A071102	Failure to Interrogate		
							E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
									A071102	Failure to Interrogate		
							F12	Serious Injury/ Illness/ Impairment	A070504	Premature Discharge of Battery		
									A071102	Failure to Interrogate		
					CT2406	Implantable	E0112	Dizziness	A070504	Premature Discharge of Battery		
									A071102	Failure to Interrogate		
							E0601	Arrhythmia	A070504	Premature Discharge of Battery		
									A071102	Failure to Interrogate		
							E0717	Dyspnea	A070504	Premature Discharge of Battery		
									A071102	Failure to Interrogate		
							E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
									A071102	Failure to Interrogate		
							F12	Serious Injury/ Illness/ Impairment	A070504	Premature Discharge of Battery		
									A071102	Failure to Interrogate		
					CT2407	Active implantable	E0112	Dizziness	A070504	Premature Discharge of Battery		
									A071102	Failure to Interrogate		
							E0601	Arrhythmia	A070504	Premature Discharge of Battery		
									A071102	Failure to Interrogate		
							E0717	Dyspnea	A070504	Premature Discharge of Battery		
									A071102	Failure to Interrogate		
							E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
									A071102	Failure to Interrogate		
							F12	Serious Injury/ Illness/ Impairment	A070504	Premature Discharge of Battery		
									A071102	Failure to Interrogate		
					CT257	Implantable cardiac pacemakers	E0112	Dizziness	A070504	Premature Discharge of Battery		
									A071102	Failure to Interrogate		
							E0601	Arrhythmia	A070504	Premature Discharge of Battery		
									A071102	Failure to Interrogate		
							E0717	Dyspnea	A070504	Premature Discharge of Battery		
									A071102	Failure to Interrogate		
							E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
									A071102	Failure to Interrogate		
							F12	Serious Injury/ Illness/ Impairment	A070504	Premature Discharge of Battery		
									A071102	Failure to Interrogate		
					CT2727	Active implant devices/ancillaries	E0112	Dizziness	A070504	Premature Discharge of Battery		
									A071102	Failure to Interrogate		
							E0601	Arrhythmia	A070504	Premature Discharge of Battery		
									A071102	Failure to Interrogate		
							E0717	Dyspnea	A070504	Premature Discharge of Battery		
									A071102	Failure to Interrogate		
							E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
									A071102	Failure to Interrogate		
							F12	Serious Injury/ Illness/ Impairment	A070504	Premature Discharge of Battery		
									A071102	Failure to Interrogate		
					CT2750	Implantable cardiac electrical stimulators and associated devices	E0112	Dizziness	A070504	Premature Discharge of Battery		
									A071102	Failure to Interrogate		
							E0601	Arrhythmia	A070504	Premature Discharge of Battery		
									A071102	Failure to Interrogate		
							E0717	Dyspnea	A070504	Premature Discharge of Battery		
									A071102	Failure to Interrogate		
							E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
									A071102	Failure to Interrogate		
							F12	Serious Injury/ Illness/ Impairment	A070504	Premature Discharge of Battery		
									A071102	Failure to Interrogate		
					CT2866	Implantable cardiac electrical stimulators	E0112	Dizziness	A070504	Premature Discharge of Battery		
									A071102	Failure to Interrogate		
							E0601	Arrhythmia	A070504	Premature Discharge of Battery		
									A071102	Failure to Interrogate		
							E0717	Dyspnea	A070504	Premature Discharge of Battery		
									A071102	Failure to Interrogate		
							E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
									A071102	Failure to Interrogate		
							F12	Serious Injury/ Illness/ Impairment	A070504	Premature Discharge of Battery		
									A071102	Failure to Interrogate		
					CT2985	Active	E0112	Dizziness	A070504	Premature Discharge of Battery		

									A071102	Failure to Interrogate		
							E0601	Arrhythmia	A070504	Premature Discharge of Battery		
									A071102	Failure to Interrogate		
							E0717	Dyspnea	A070504	Premature Discharge of Battery		
									A071102	Failure to Interrogate		
							E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
									A071102	Failure to Interrogate		
							F12	Serious Injury/ Illness/ Impairment	A070504	Premature Discharge of Battery		
									A071102	Failure to Interrogate		
					CT3001	Implantable metal	E0112	Dizziness	A070504	Premature Discharge of Battery		
									A071102	Failure to Interrogate		
							E0601	Arrhythmia	A070504	Premature Discharge of Battery		
									A071102	Failure to Interrogate		
							E0717	Dyspnea	A070504	Premature Discharge of Battery		
									A071102	Failure to Interrogate		
							E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
									A071102	Failure to Interrogate		
							F12	Serious Injury/ Illness/ Impairment	A070504	Premature Discharge of Battery		
									A071102	Failure to Interrogate		
					CT3002	Implantable synthetic polymer	E0112	Dizziness	A070504	Premature Discharge of Battery		
									A071102	Failure to Interrogate		
							E0601	Arrhythmia	A070504	Premature Discharge of Battery		
									A071102	Failure to Interrogate		
							E0717	Dyspnea	A070504	Premature Discharge of Battery		
									A071102	Failure to Interrogate		
							E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
									A071102	Failure to Interrogate		
							F12	Serious Injury/ Illness/ Impairment	A070504	Premature Discharge of Battery		
									A071102	Failure to Interrogate		
					CT3006	Implantable inorganic material	E0112	Dizziness	A070504	Premature Discharge of Battery		
									A071102	Failure to Interrogate		
							E0601	Arrhythmia	A070504	Premature Discharge of Battery		
									A071102	Failure to Interrogate		
							E0717	Dyspnea	A070504	Premature Discharge of Battery		
									A071102	Failure to Interrogate		
							E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
									A071102	Failure to Interrogate		
							F12	Serious Injury/ Illness/ Impairment	A070504	Premature Discharge of Battery		
									A071102	Failure to Interrogate		
					CT314	Electrically powered	E0112	Dizziness	A070504	Premature Discharge of Battery		
									A071102	Failure to Interrogate		
							E0601	Arrhythmia	A070504	Premature Discharge of Battery		
									A071102	Failure to Interrogate		
							E0717	Dyspnea	A070504	Premature Discharge of Battery		
									A071102	Failure to Interrogate		
							E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
									A071102	Failure to Interrogate		
							F12	Serious Injury/ Illness/ Impairment	A070504	Premature Discharge of Battery		
									A071102	Failure to Interrogate		
					CT321	Long-term surgical invasive	E0112	Dizziness	A070504	Premature Discharge of Battery		
									A071102	Failure to Interrogate		
							E0601	Arrhythmia	A070504	Premature Discharge of Battery		
									A071102	Failure to Interrogate		
							E0717	Dyspnea	A070504	Premature Discharge of Battery		
									A071102	Failure to Interrogate		
							E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
									A071102	Failure to Interrogate		
							F12	Serious Injury/ Illness/ Impairment	A070504	Premature Discharge of Battery		
									A071102	Failure to Interrogate		
					CT449	Generators and associated devices	E0112	Dizziness	A070504	Premature Discharge of Battery		
									A071102	Failure to Interrogate		
							E0601	Arrhythmia	A070504	Premature Discharge of Battery		
									A071102	Failure to Interrogate		
							E0717	Dyspnea	A070504	Premature Discharge of Battery		
									A071102	Failure to Interrogate		
							E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
									A071102	Failure to Interrogate		
							F12	Serious Injury/ Illness/ Impairment	A070504	Premature Discharge of Battery		
									A071102	Failure to Interrogate		
					CT981	Single use	E0112	Dizziness	A070504	Premature Discharge of Battery		
									A071102	Failure to Interrogate		
							E0601	Arrhythmia	A070504	Premature Discharge of Battery		
									A071102	Failure to Interrogate		
							E0717	Dyspnea	A070504	Premature Discharge of Battery		
									A071102	Failure to Interrogate		
							E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
									A071102	Failure to Interrogate		
							F12	Serious Injury/ Illness/ Impairment	A070504	Premature Discharge of Battery		
									A071102	Failure to Interrogate		
					CT983	Surgical invasive	E0112	Dizziness	A070504	Premature Discharge of Battery		
									A071102	Failure to Interrogate		
							E0601	Arrhythmia	A070504	Premature Discharge of Battery		
									A071102	Failure to Interrogate		
							E0717	Dyspnea	A070504	Premature Discharge of Battery		
									A071102	Failure to Interrogate		
							E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
									A071102	Failure to Interrogate		
							F12	Serious Injury/ Illness/ Impairment	A070504	Premature Discharge of Battery		

									A071102	Failure to Interrogate		
7015	2021	All other reportable incidents	47265	Dual-chamber implantable pacemaker, rate-responsive	CT141	Electrodes/Leads and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	23/11/2021	23/11/2021
							F1901	Additional Surgery	A070504	Premature Discharge of Battery	23/11/2021	23/11/2021
					CT145	Cardiovascular devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	23/11/2021	23/11/2021
							F1901	Additional Surgery	A070504	Premature Discharge of Battery	23/11/2021	23/11/2021
					CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	23/11/2021	23/11/2021
							F1901	Additional Surgery	A070504	Premature Discharge of Battery	23/11/2021	23/11/2021
					CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	23/11/2021	23/11/2021
							F1901	Additional Surgery	A070504	Premature Discharge of Battery	23/11/2021	23/11/2021
					CT233	Surgical	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	23/11/2021	23/11/2021
							F1901	Additional Surgery	A070504	Premature Discharge of Battery	23/11/2021	23/11/2021
					CT2406	Implantable	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	23/11/2021	23/11/2021
							F1901	Additional Surgery	A070504	Premature Discharge of Battery	23/11/2021	23/11/2021
					CT2407	Active implantable	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	23/11/2021	23/11/2021
							F1901	Additional Surgery	A070504	Premature Discharge of Battery	23/11/2021	23/11/2021
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	23/11/2021	23/11/2021
							F1901	Additional Surgery	A070504	Premature Discharge of Battery	23/11/2021	23/11/2021
					CT2727	Active implant devices/ancillaries	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	23/11/2021	23/11/2021
							F1901	Additional Surgery	A070504	Premature Discharge of Battery	23/11/2021	23/11/2021
					CT2750	Implantable cardiac electrical stimulators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	23/11/2021	23/11/2021
							F1901	Additional Surgery	A070504	Premature Discharge of Battery	23/11/2021	23/11/2021
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	23/11/2021	23/11/2021
							F1901	Additional Surgery	A070504	Premature Discharge of Battery	23/11/2021	23/11/2021
					CT2985	Active	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	23/11/2021	23/11/2021
							F1901	Additional Surgery	A070504	Premature Discharge of Battery	23/11/2021	23/11/2021
					CT3001	Implantable metal	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	23/11/2021	23/11/2021
							F1901	Additional Surgery	A070504	Premature Discharge of Battery	23/11/2021	23/11/2021
					CT3002	Implantable synthetic polymer	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	23/11/2021	23/11/2021
							F1901	Additional Surgery	A070504	Premature Discharge of Battery	23/11/2021	23/11/2021
					CT3006	Implantable inorganic material	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	23/11/2021	23/11/2021
							F1901	Additional Surgery	A070504	Premature Discharge of Battery	23/11/2021	23/11/2021
					CT314	Electrically powered	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	23/11/2021	23/11/2021
							F1901	Additional Surgery	A070504	Premature Discharge of Battery	23/11/2021	23/11/2021
					CT321	Long-term surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	23/11/2021	23/11/2021
							F1901	Additional Surgery	A070504	Premature Discharge of Battery	23/11/2021	23/11/2021
					CT449	Generators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	23/11/2021	23/11/2021
							F1901	Additional Surgery	A070504	Premature Discharge of Battery	23/11/2021	23/11/2021
					CT981	Single use	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	23/11/2021	23/11/2021
							F1901	Additional Surgery	A070504	Premature Discharge of Battery	23/11/2021	23/11/2021
					CT983	Surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	23/11/2021	23/11/2021
							F1901	Additional Surgery	A070504	Premature Discharge of Battery	23/11/2021	23/11/2021
7016	2021	All other reportable incidents	35852	Single-chamber implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A0508	Noise, Audible		
									A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A0508	Noise, Audible		
									A070504	Premature Discharge of Battery		
7017	2021	All other reportable incidents	37265	Dual-chamber implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A0908	Incorrect, Inadequate or Imprecise Result or Readings	27/10/2021	27/10/2021
							F08	Hospitalization or Prolonged Hospitalization	A0908	Incorrect, Inadequate or Imprecise Result or Readings	27/10/2021	27/10/2021
							F1903	Device Explantation	A0908	Incorrect, Inadequate or Imprecise Result or Readings	27/10/2021	27/10/2021
7018	2021	All other reportable incidents	47265	Dual-chamber implantable pacemaker, rate-responsive	CT141	Electrodes/Leads and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	28/10/2021	28/10/2021
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	28/10/2021	28/10/2021
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	28/10/2021	28/10/2021
					CT145	Cardiovascular devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	28/10/2021	28/10/2021
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	28/10/2021	28/10/2021
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	28/10/2021	28/10/2021
					CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	28/10/2021	28/10/2021
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	28/10/2021	28/10/2021
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	28/10/2021	28/10/2021
					CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	28/10/2021	28/10/2021
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	28/10/2021	28/10/2021
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	28/10/2021	28/10/2021
					CT233	Surgical	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	28/10/2021	28/10/2021
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	28/10/2021	28/10/2021
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	28/10/2021	28/10/2021
					CT2406	Implantable	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	28/10/2021	28/10/2021
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	28/10/2021	28/10/2021
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	28/10/2021	28/10/2021
					CT2407	Active implantable	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	28/10/2021	28/10/2021
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	28/10/2021	28/10/2021
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	28/10/2021	28/10/2021
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	28/10/2021	28/10/2021
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	28/10/2021	28/10/2021
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	28/10/2021	28/10/2021
					CT2727	Active implant devices/ancillaries	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	28/10/2021	28/10/2021
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	28/10/2021	28/10/2021
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	28/10/2021	28/10/2021
					CT2750	Implantable cardiac electrical stimulators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	28/10/2021	28/10/2021
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	28/10/2021	28/10/2021
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	28/10/2021	28/10/2021
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	28/10/2021	28/10/2021
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	28/10/2021	28/10/2021
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	28/10/2021	28/10/2021
					CT2985	Active	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	28/10/2021	28/10/2021
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	28/10/2021	28/10/2021
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	28/10/2021	28/10/2021
					CT3001	Implantable metal	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	28/10/2021	28/10/2021

							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	28/10/2021	28/10/2021
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	28/10/2021	28/10/2021
					CT3002	Implantable synthetic polymer	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	28/10/2021	28/10/2021
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	28/10/2021	28/10/2021
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	28/10/2021	28/10/2021
					CT3006	Implantable inorganic material	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	28/10/2021	28/10/2021
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	28/10/2021	28/10/2021
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	28/10/2021	28/10/2021
					CT314	Electrically powered	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	28/10/2021	28/10/2021
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	28/10/2021	28/10/2021
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	28/10/2021	28/10/2021
					CT321	Long-term surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	28/10/2021	28/10/2021
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	28/10/2021	28/10/2021
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	28/10/2021	28/10/2021
					CT449	Generators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	28/10/2021	28/10/2021
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	28/10/2021	28/10/2021
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	28/10/2021	28/10/2021
					CT981	Single use	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	28/10/2021	28/10/2021
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	28/10/2021	28/10/2021
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	28/10/2021	28/10/2021
					CT983	Surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	28/10/2021	28/10/2021
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	28/10/2021	28/10/2021
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	28/10/2021	28/10/2021
7019	2021	All other reportable incidents	47267	Single-chamber implantable pacemaker, rate-responsive	CT141	Electrodes/Leads and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT145	Cardiovascular devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT233	Surgical	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT2406	Implantable	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT2407	Active implantable	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT2727	Active implant devices/ancillaries	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT2750	Implantable cardiac electrical stimulators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT2985	Active	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT3001	Implantable metal	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT3002	Implantable synthetic polymer	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT3006	Implantable inorganic material	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT314	Electrically powered	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT321	Long-term surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT449	Generators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT981	Single use	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT983	Surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
7020	2022	All other reportable incidents	37265	Dual-chamber implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	17/06/2022	17/06/2022
							F1901	Additional Surgery	A070504	Premature Discharge of Battery	17/06/2022	17/06/2022
7021	2022	All other reportable incidents	47205	Cardiac pulse generator programmer	CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A090805	Incorrect Measurement		
							F23	Unexpected Medical Intervention	A090805	Incorrect Measurement		
7022	2022	All other reportable incidents	35223	Endocardial/interventricular septal pacing lead	CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A051201	Device Dislodged or Dislocated	17/06/2022	17/06/2022
							F1901	Additional Surgery	A051201	Device Dislodged or Dislocated	17/06/2022	17/06/2022
7023	2021	All other reportable incidents	35223	Endocardial/interventricular septal pacing lead	CT1684	Cardiac pacemakers and associated devices	E200801	Device Embedded In Tissue or Plaque	A040101	Fracture	11/10/2021	11/10/2021
							E2403	No Clinical Signs, Symptoms or Conditions	A040101	Fracture	11/10/2021	11/10/2021
							F12	Serious Injury/ Illness/ Impairment	A040101	Fracture	11/10/2021	11/10/2021
7024	2022	All other reportable incidents	35223	Endocardial/interventricular septal pacing lead	CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070909	Oversensing		
									A0712	Pacing Problem		
									A090801	Signal Artifact/Noise		
							F11	Non-Serious Injury/ Illness/ Impairment	A070909	Oversensing		
									A0712	Pacing Problem		
									A090801	Signal Artifact/Noise		
							F23	Unexpected Medical Intervention	A070909	Oversensing		
									A0712	Pacing Problem		
									A090801	Signal Artifact/Noise		
7025	2022	All other reportable incidents	35223	Endocardial/interventricular septal pacing lead	CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070909	Oversensing		
									A090801	Signal Artifact/Noise		
							F26	No Health Consequences or Impact	A070909	Oversensing		
									A090801	Signal Artifact/Noise		
7026	2022	All other reportable incidents	35223	Endocardial/interventricular septal pacing lead	CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070909	Oversensing		
									A090801	Signal Artifact/Noise		
							F26	No Health Consequences or Impact	A070909	Oversensing		

									A090801	Signal Artifact/Noise		
7027	2022	All other reportable incidents	47265	Dual-chamber implantable pacemaker, rate-responsive	CT141	Electrodes/Leads and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	23/06/2022	23/06/2022
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	23/06/2022	23/06/2022
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	23/06/2022	23/06/2022
					CT145	Cardiovascular devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	23/06/2022	23/06/2022
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	23/06/2022	23/06/2022
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	23/06/2022	23/06/2022
					CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	23/06/2022	23/06/2022
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	23/06/2022	23/06/2022
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	23/06/2022	23/06/2022
					CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	23/06/2022	23/06/2022
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	23/06/2022	23/06/2022
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	23/06/2022	23/06/2022
					CT233	Surgical	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	23/06/2022	23/06/2022
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	23/06/2022	23/06/2022
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	23/06/2022	23/06/2022
					CT2406	Implantable	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	23/06/2022	23/06/2022
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	23/06/2022	23/06/2022
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	23/06/2022	23/06/2022
					CT2407	Active implantable	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	23/06/2022	23/06/2022
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	23/06/2022	23/06/2022
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	23/06/2022	23/06/2022
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	23/06/2022	23/06/2022
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	23/06/2022	23/06/2022
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	23/06/2022	23/06/2022
					CT2727	Active implant devices/ancillaries	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	23/06/2022	23/06/2022
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	23/06/2022	23/06/2022
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	23/06/2022	23/06/2022
					CT2750	Implantable cardiac electrical stimulators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	23/06/2022	23/06/2022
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	23/06/2022	23/06/2022
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	23/06/2022	23/06/2022
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	23/06/2022	23/06/2022
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	23/06/2022	23/06/2022
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	23/06/2022	23/06/2022
					CT2985	Active	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	23/06/2022	23/06/2022
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	23/06/2022	23/06/2022
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	23/06/2022	23/06/2022
					CT3001	Implantable metal	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	23/06/2022	23/06/2022
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	23/06/2022	23/06/2022
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	23/06/2022	23/06/2022
					CT3002	Implantable synthetic polymer	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	23/06/2022	23/06/2022
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	23/06/2022	23/06/2022
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	23/06/2022	23/06/2022
					CT3006	Implantable inorganic material	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	23/06/2022	23/06/2022
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	23/06/2022	23/06/2022
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	23/06/2022	23/06/2022
					CT314	Electrically powered	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	23/06/2022	23/06/2022
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	23/06/2022	23/06/2022
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	23/06/2022	23/06/2022
					CT321	Long-term surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	23/06/2022	23/06/2022
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	23/06/2022	23/06/2022
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	23/06/2022	23/06/2022
					CT449	Generators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	23/06/2022	23/06/2022
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	23/06/2022	23/06/2022
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	23/06/2022	23/06/2022
					CT981	Single use	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	23/06/2022	23/06/2022
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	23/06/2022	23/06/2022
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	23/06/2022	23/06/2022
					CT983	Surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	23/06/2022	23/06/2022
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	23/06/2022	23/06/2022
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	23/06/2022	23/06/2022
7028	2022	All other reportable incidents	35852	Single-chamber implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem	06/10/2022	06/10/2022
									A070504	Premature Discharge of Battery	06/10/2022	06/10/2022
							F08	Hospitalization or Prolonged Hospitalization	A0705	Battery Problem	06/10/2022	06/10/2022
									A070504	Premature Discharge of Battery	06/10/2022	06/10/2022
							F1905	Device Revision or Replacement	A0705	Battery Problem	06/10/2022	06/10/2022
									A070504	Premature Discharge of Battery	06/10/2022	06/10/2022
7029	2022	All other reportable incidents	17882	Physiologic-monitoring defibrillation system	CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070803	Failure to Power Up		
							F27	Problem Identified During Non-Clinical Procedure	A070803	Failure to Power Up		
7030	2022	All other reportable incidents	17882	Physiologic-monitoring defibrillation system	CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A071301	Failure to Deliver Shock/Stimulation		
							F27	Problem Identified During Non-Clinical Procedure	A071301	Failure to Deliver Shock/Stimulation		
7031	2022	All other reportable incidents	17882	Physiologic-monitoring defibrillation system	CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070803	Failure to Power Up		
							F27	Problem Identified During Non-Clinical Procedure	A070803	Failure to Power Up		
7032	2021	All other reportable incidents	47270	Cardiac resynchronization therapy implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	22/10/2021	22/10/2021
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery	22/10/2021	22/10/2021
7033	2022	All other reportable incidents	17882	Physiologic-monitoring defibrillation system	CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070803	Failure to Power Up		
							F27	Problem Identified During Non-Clinical Procedure	A070803	Failure to Power Up		
7034	2022	All other reportable incidents	17882	Physiologic-monitoring defibrillation system	CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070803	Failure to Power Up		
							F27	Problem Identified During Non-Clinical Procedure	A070803	Failure to Power Up		
7035	2021	All other reportable incidents	35852	Single-chamber implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A070909	Oversensing		
									A071302	Inappropriate/Inadequate Shock/Stimulation		
									A1301	Failure to Read Input Signal		
							F26	No Health Consequences or Impact	A070909	Oversensing		
									A071302	Inappropriate/Inadequate Shock/Stimulation		
									A1301	Failure to Read Input Signal		
7036	2021	All other reportable incidents	35223	Endocardial/interventricular septal pacing lead	CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A0401	Break		
							F26	No Health Consequences or Impact	A0401	Break		
7037	2021	All other reportable incidents	37265	Dual-chamber implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	01/11/2021	01/11/2021

									A090805	Incorrect Measurement	01/11/2021	01/11/2021
							F1901	Additional Surgery	A070504	Premature Discharge of Battery	01/11/2021	01/11/2021
									A090805	Incorrect Measurement	01/11/2021	01/11/2021
7038	2021	All other reportable incidents	35223	Endocardial/interventricular septal pacing lead	CT1684	Cardiac pacemakers and associated devices	E060104	Bradycardia	A070101	Failure to Capture		
									A070102	High Capture Threshold		
									A0722	Impedance Problem		
							E2403	No Clinical Signs, Symptoms or Conditions	A070101	Failure to Capture		
									A070102	High Capture Threshold		
									A0722	Impedance Problem		
							F08	Hospitalization or Prolonged Hospitalization	A070101	Failure to Capture		
									A070102	High Capture Threshold		
									A0722	Impedance Problem		
							F11	Non-Serious Injury/ Illness/ Impairment	A070101	Failure to Capture		
									A070102	High Capture Threshold		
									A0722	Impedance Problem		
							F23	Unexpected Medical Intervention	A070101	Failure to Capture		
									A070102	High Capture Threshold		
									A0722	Impedance Problem		
7039	2021	All other reportable incidents	47263	Cardiac resynchronization therapy implantable pacemaker	CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	05/11/2021	05/11/2021
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	05/11/2021	05/11/2021
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	05/11/2021	05/11/2021
					CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	05/11/2021	05/11/2021
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	05/11/2021	05/11/2021
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	05/11/2021	05/11/2021
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	05/11/2021	05/11/2021
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	05/11/2021	05/11/2021
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	05/11/2021	05/11/2021
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	05/11/2021	05/11/2021
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	05/11/2021	05/11/2021
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	05/11/2021	05/11/2021
7040	2021	All other reportable incidents	35223	Endocardial/interventricular septal pacing lead	CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A0401	Break	11/10/2021	
									A040101	Fracture	11/10/2021	
							F26	No Health Consequences or Impact	A0401	Break	11/10/2021	
									A040101	Fracture	11/10/2021	
7041	2021	All other reportable incidents	35852	Single-chamber implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem	26/06/2022	26/06/2022
									A070504	Premature Discharge of Battery	26/06/2022	26/06/2022
							F08	Hospitalization or Prolonged Hospitalization	A0705	Battery Problem	26/06/2022	26/06/2022
									A070504	Premature Discharge of Battery	26/06/2022	26/06/2022
							F1903	Device Explantation	A0705	Battery Problem	26/06/2022	26/06/2022
									A070504	Premature Discharge of Battery	26/06/2022	26/06/2022
7042	2021	All other reportable incidents	35223	Endocardial/interventricular septal pacing lead	CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A040101	Fracture	04/11/2021	04/11/2021
									A070909	Oversensing	04/11/2021	04/11/2021
									A072201	High Impedance	04/11/2021	04/11/2021
									A090801	Signal Artifact/Noise	04/11/2021	04/11/2021
							F26	No Health Consequences or Impact	A040101	Fracture	04/11/2021	04/11/2021
									A070909	Oversensing	04/11/2021	04/11/2021
									A072201	High Impedance	04/11/2021	04/11/2021
									A090801	Signal Artifact/Noise	04/11/2021	04/11/2021
7043	2021	All other reportable incidents	35223	Endocardial/interventricular septal pacing lead	CT1684	Cardiac pacemakers and associated devices	E060106	Heart Block	A05	Mechanical Problem		
									A0701	Capturing Problem		
									A0722	Impedance Problem		
									A072202	Low Impedance		
							E2403	No Clinical Signs, Symptoms or Conditions	A05	Mechanical Problem		
									A0701	Capturing Problem		
									A0722	Impedance Problem		
									A072202	Low Impedance		
							F12	Serious Injury/ Illness/ Impairment	A05	Mechanical Problem		
									A0701	Capturing Problem		
									A0722	Impedance Problem		
									A072202	Low Impedance		
7044	2021	All other reportable incidents	17882	Physiologic-monitoring defibrillation system	CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A0719	Unexpected Shutdown		
							F26	No Health Consequences or Impact	A0719	Unexpected Shutdown		
7045	2021	All other reportable incidents	17882	Physiologic-monitoring defibrillation system	CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070803	Failure to Power Up		
							F26	No Health Consequences or Impact	A070803	Failure to Power Up		
7046	2021	All other reportable incidents	35223	Endocardial/interventricular septal pacing lead	CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A04	Material Integrity Problem		
									A070909	Oversensing		
									A0722	Impedance Problem		
									A090801	Signal Artifact/Noise		
							F26	No Health Consequences or Impact	A04	Material Integrity Problem		
									A070909	Oversensing		
									A0722	Impedance Problem		
									A090801	Signal Artifact/Noise		
7047	2021	All other reportable incidents	60789	Intracardiac pacemaker	CT130	Catheters and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A1502	Positioning Problem	31/08/2021	31/08/2021
							F26	No Health Consequences or Impact	A1502	Positioning Problem	31/08/2021	31/08/2021
					CT145	Cardiovascular devices	E2403	No Clinical Signs, Symptoms or Conditions	A1502	Positioning Problem	31/08/2021	31/08/2021
							F26	No Health Consequences or Impact	A1502	Positioning Problem	31/08/2021	31/08/2021
					CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A1502	Positioning Problem	31/08/2021	31/08/2021
							F26	No Health Consequences or Impact	A1502	Positioning Problem	31/08/2021	31/08/2021
					CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A1502	Positioning Problem	31/08/2021	31/08/2021
							F26	No Health Consequences or Impact	A1502	Positioning Problem	31/08/2021	31/08/2021
					CT233	Surgical	E2403	No Clinical Signs, Symptoms or Conditions	A1502	Positioning Problem	31/08/2021	31/08/2021
							F26	No Health Consequences or Impact	A1502	Positioning Problem	31/08/2021	31/08/2021
					CT2406	Implantable	E2403	No Clinical Signs, Symptoms or Conditions	A1502	Positioning Problem	31/08/2021	31/08/2021
							F26	No Health Consequences or Impact	A1502	Positioning Problem	31/08/2021	31/08/2021
					CT2407	Active implantable	E2403	No Clinical Signs, Symptoms or Conditions	A1502	Positioning Problem	31/08/2021	31/08/2021
							F26	No Health Consequences or Impact	A1502	Positioning Problem	31/08/2021	31/08/2021
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A1502	Positioning Problem	31/08/2021	31/08/2021
							F26	No Health Consequences or Impact	A1502	Positioning Problem	31/08/2021	31/08/2021

					CT2727	Active implant devices/ancillaries	E2403	No Clinical Signs, Symptoms or Conditions	A1502	Positioning Problem	31/08/2021	31/08/2021
							F26	No Health Consequences or Impact	A1502	Positioning Problem	31/08/2021	31/08/2021
					CT2750	Implantable cardiac electrical stimulators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A1502	Positioning Problem	31/08/2021	31/08/2021
							F26	No Health Consequences or Impact	A1502	Positioning Problem	31/08/2021	31/08/2021
					CT2818	Catheterization/Cannulation/Fluid path devices	E2403	No Clinical Signs, Symptoms or Conditions	A1502	Positioning Problem	31/08/2021	31/08/2021
							F26	No Health Consequences or Impact	A1502	Positioning Problem	31/08/2021	31/08/2021
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A1502	Positioning Problem	31/08/2021	31/08/2021
							F26	No Health Consequences or Impact	A1502	Positioning Problem	31/08/2021	31/08/2021
					CT2985	Active	E2403	No Clinical Signs, Symptoms or Conditions	A1502	Positioning Problem	31/08/2021	31/08/2021
							F26	No Health Consequences or Impact	A1502	Positioning Problem	31/08/2021	31/08/2021
					CT3001	Implantable metal	E2403	No Clinical Signs, Symptoms or Conditions	A1502	Positioning Problem	31/08/2021	31/08/2021
							F26	No Health Consequences or Impact	A1502	Positioning Problem	31/08/2021	31/08/2021
					CT3002	Implantable synthetic polymer	E2403	No Clinical Signs, Symptoms or Conditions	A1502	Positioning Problem	31/08/2021	31/08/2021
							F26	No Health Consequences or Impact	A1502	Positioning Problem	31/08/2021	31/08/2021
					CT3006	Implantable inorganic material	E2403	No Clinical Signs, Symptoms or Conditions	A1502	Positioning Problem	31/08/2021	31/08/2021
							F26	No Health Consequences or Impact	A1502	Positioning Problem	31/08/2021	31/08/2021
					CT314	Electrically powered	E2403	No Clinical Signs, Symptoms or Conditions	A1502	Positioning Problem	31/08/2021	31/08/2021
							F26	No Health Consequences or Impact	A1502	Positioning Problem	31/08/2021	31/08/2021
					CT321	Long-term surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A1502	Positioning Problem	31/08/2021	31/08/2021
							F26	No Health Consequences or Impact	A1502	Positioning Problem	31/08/2021	31/08/2021
					CT981	Single use	E2403	No Clinical Signs, Symptoms or Conditions	A1502	Positioning Problem	31/08/2021	31/08/2021
							F26	No Health Consequences or Impact	A1502	Positioning Problem	31/08/2021	31/08/2021
					CT983	Surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A1502	Positioning Problem	31/08/2021	31/08/2021
							F26	No Health Consequences or Impact	A1502	Positioning Problem	31/08/2021	31/08/2021
7048	2021	All other reportable incidents	17882	Physiologic-monitoring defibrillation system	CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A071301	Failure to Deliver Shock/Stimulation		
							F27	Problem Identified During Non-Clinical Procedure	A071301	Failure to Deliver Shock/Stimulation		
7049	2022	All other reportable incidents	47267	Single-chamber implantable pacemaker, rate-responsive	CT141	Electrodes/Leads and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	06/09/2022	06/09/2022
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	06/09/2022	06/09/2022
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	06/09/2022	06/09/2022
					CT145	Cardiovascular devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	06/09/2022	06/09/2022
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	06/09/2022	06/09/2022
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	06/09/2022	06/09/2022
					CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	06/09/2022	06/09/2022
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	06/09/2022	06/09/2022
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	06/09/2022	06/09/2022
					CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	06/09/2022	06/09/2022
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	06/09/2022	06/09/2022
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	06/09/2022	06/09/2022
					CT233	Surgical	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	06/09/2022	06/09/2022
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	06/09/2022	06/09/2022
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	06/09/2022	06/09/2022
					CT2406	Implantable	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	06/09/2022	06/09/2022
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	06/09/2022	06/09/2022
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	06/09/2022	06/09/2022
					CT2407	Active implantable	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	06/09/2022	06/09/2022
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	06/09/2022	06/09/2022
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	06/09/2022	06/09/2022
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	06/09/2022	06/09/2022
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	06/09/2022	06/09/2022
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	06/09/2022	06/09/2022
					CT2727	Active implant devices/ancillaries	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	06/09/2022	06/09/2022
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	06/09/2022	06/09/2022
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	06/09/2022	06/09/2022
					CT2750	Implantable cardiac electrical stimulators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	06/09/2022	06/09/2022
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	06/09/2022	06/09/2022
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	06/09/2022	06/09/2022
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	06/09/2022	06/09/2022
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	06/09/2022	06/09/2022
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	06/09/2022	06/09/2022
					CT2985	Active	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	06/09/2022	06/09/2022
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	06/09/2022	06/09/2022
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	06/09/2022	06/09/2022
					CT3001	Implantable metal	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	06/09/2022	06/09/2022
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	06/09/2022	06/09/2022
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	06/09/2022	06/09/2022
					CT3002	Implantable synthetic polymer	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	06/09/2022	06/09/2022
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	06/09/2022	06/09/2022
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	06/09/2022	06/09/2022
					CT3006	Implantable inorganic material	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	06/09/2022	06/09/2022
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	06/09/2022	06/09/2022
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	06/09/2022	06/09/2022
					CT314	Electrically powered	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	06/09/2022	06/09/2022
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	06/09/2022	06/09/2022
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	06/09/2022	06/09/2022
					CT321	Long-term surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	06/09/2022	06/09/2022
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	06/09/2022	06/09/2022
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	06/09/2022	06/09/2022
					CT449	Generators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	06/09/2022	06/09/2022
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	06/09/2022	06/09/2022
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	06/09/2022	06/09/2022
					CT981	Single use	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	06/09/2022	06/09/2022
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	06/09/2022	06/09/2022
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	06/09/2022	06/09/2022
					CT983	Surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	06/09/2022	06/09/2022
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	06/09/2022	06/09/2022
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	06/09/2022	06/09/2022
7050	2022	All other reportable incidents	35852	Single-chamber implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	22/09/2022	22/09/2022
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	22/09/2022	22/09/2022

							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	22/09/2022	22/09/2022
7051	2022	All other reportable incidents	17882	Physiologic-monitoring defibrillation system	CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A0719	Unexpected Shutdown		
							F27	Problem Identified During Non-Clinical Procedure	A0719	Unexpected Shutdown		
7052	2022	All other reportable incidents	47263	Cardiac resynchronization therapy implantable pacemaker	CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem	05/08/2022	05/08/2022
									A070504	Premature Discharge of Battery	05/08/2022	05/08/2022
							F08	Hospitalization or Prolonged Hospitalization	A0705	Battery Problem	05/08/2022	05/08/2022
									A070504	Premature Discharge of Battery	05/08/2022	05/08/2022
							F1905	Device Revision or Replacement	A0705	Battery Problem	05/08/2022	05/08/2022
									A070504	Premature Discharge of Battery	05/08/2022	05/08/2022
					CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem	05/08/2022	05/08/2022
									A070504	Premature Discharge of Battery	05/08/2022	05/08/2022
							F08	Hospitalization or Prolonged Hospitalization	A0705	Battery Problem	05/08/2022	05/08/2022
									A070504	Premature Discharge of Battery	05/08/2022	05/08/2022
							F1905	Device Revision or Replacement	A0705	Battery Problem	05/08/2022	05/08/2022
									A070504	Premature Discharge of Battery	05/08/2022	05/08/2022
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem	05/08/2022	05/08/2022
									A070504	Premature Discharge of Battery	05/08/2022	05/08/2022
							F08	Hospitalization or Prolonged Hospitalization	A0705	Battery Problem	05/08/2022	05/08/2022
									A070504	Premature Discharge of Battery	05/08/2022	05/08/2022
							F1905	Device Revision or Replacement	A0705	Battery Problem	05/08/2022	05/08/2022
									A070504	Premature Discharge of Battery	05/08/2022	05/08/2022
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem	05/08/2022	05/08/2022
									A070504	Premature Discharge of Battery	05/08/2022	05/08/2022
							F08	Hospitalization or Prolonged Hospitalization	A0705	Battery Problem	05/08/2022	05/08/2022
									A070504	Premature Discharge of Battery	05/08/2022	05/08/2022
							F1905	Device Revision or Replacement	A0705	Battery Problem	05/08/2022	05/08/2022
									A070504	Premature Discharge of Battery	05/08/2022	05/08/2022
7053	2022	All other reportable incidents	17882	Physiologic-monitoring defibrillation system	CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070908	Failure to Sense		
							F27	Problem Identified During Non-Clinical Procedure	A070908	Failure to Sense		
7054	2021	All other reportable incidents	36241	Pacemaker/defibrillator lead	CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A072201	High Impedance		
							F26	No Health Consequences or Impact	A072201	High Impedance		
7055	2022	All other reportable incidents	35852	Single-chamber implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem	08/08/2022	08/08/2022
									A070504	Premature Discharge of Battery	08/08/2022	08/08/2022
							F08	Hospitalization or Prolonged Hospitalization	A0705	Battery Problem	08/08/2022	08/08/2022
									A070504	Premature Discharge of Battery	08/08/2022	08/08/2022
							F1905	Device Revision or Replacement	A0705	Battery Problem	08/08/2022	08/08/2022
									A070504	Premature Discharge of Battery	08/08/2022	08/08/2022
7056	2021	All other reportable incidents	35852	Single-chamber implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem	25/01/2022	25/01/2022
									A070504	Premature Discharge of Battery	25/01/2022	25/01/2022
							F08	Hospitalization or Prolonged Hospitalization	A0705	Battery Problem	25/01/2022	25/01/2022
									A070504	Premature Discharge of Battery	25/01/2022	25/01/2022
							F1905	Device Revision or Replacement	A0705	Battery Problem	25/01/2022	25/01/2022
									A070504	Premature Discharge of Battery	25/01/2022	25/01/2022
7057	2022	All other reportable incidents	47263	Cardiac resynchronization therapy implantable pacemaker	CT1684	Cardiac pacemakers and associated devices	E0112	Dizziness	A1107	Data Problem	03/06/2022	03/06/2022
									A160503	Inappropriate or Unexpected Reset	03/06/2022	03/06/2022
							E012207	Twitching	A1107	Data Problem	03/06/2022	03/06/2022
									A160503	Inappropriate or Unexpected Reset	03/06/2022	03/06/2022
							E2403	No Clinical Signs, Symptoms or Conditions	A1107	Data Problem	03/06/2022	03/06/2022
									A160503	Inappropriate or Unexpected Reset	03/06/2022	03/06/2022
							F08	Hospitalization or Prolonged Hospitalization	A1107	Data Problem	03/06/2022	03/06/2022
									A160503	Inappropriate or Unexpected Reset	03/06/2022	03/06/2022
							F1905	Device Revision or Replacement	A1107	Data Problem	03/06/2022	03/06/2022
									A160503	Inappropriate or Unexpected Reset	03/06/2022	03/06/2022
					CT1685	Cardiac pacemakers	E0112	Dizziness	A1107	Data Problem	03/06/2022	03/06/2022
									A160503	Inappropriate or Unexpected Reset	03/06/2022	03/06/2022
							E012207	Twitching	A1107	Data Problem	03/06/2022	03/06/2022
									A160503	Inappropriate or Unexpected Reset	03/06/2022	03/06/2022
							E2403	No Clinical Signs, Symptoms or Conditions	A1107	Data Problem	03/06/2022	03/06/2022
									A160503	Inappropriate or Unexpected Reset	03/06/2022	03/06/2022
							F08	Hospitalization or Prolonged Hospitalization	A1107	Data Problem	03/06/2022	03/06/2022
									A160503	Inappropriate or Unexpected Reset	03/06/2022	03/06/2022
							F1905	Device Revision or Replacement	A1107	Data Problem	03/06/2022	03/06/2022
									A160503	Inappropriate or Unexpected Reset	03/06/2022	03/06/2022
					CT257	Implantable cardiac pacemakers	E0112	Dizziness	A1107	Data Problem	03/06/2022	03/06/2022
									A160503	Inappropriate or Unexpected Reset	03/06/2022	03/06/2022
							E012207	Twitching	A1107	Data Problem	03/06/2022	03/06/2022
									A160503	Inappropriate or Unexpected Reset	03/06/2022	03/06/2022
							E2403	No Clinical Signs, Symptoms or Conditions	A1107	Data Problem	03/06/2022	03/06/2022
									A160503	Inappropriate or Unexpected Reset	03/06/2022	03/06/2022
							F08	Hospitalization or Prolonged Hospitalization	A1107	Data Problem	03/06/2022	03/06/2022
									A160503	Inappropriate or Unexpected Reset	03/06/2022	03/06/2022
							F1905	Device Revision or Replacement	A1107	Data Problem	03/06/2022	03/06/2022
									A160503	Inappropriate or Unexpected Reset	03/06/2022	03/06/2022
					CT2866	Implantable cardiac electrical stimulators	E0112	Dizziness	A1107	Data Problem	03/06/2022	03/06/2022
									A160503	Inappropriate or Unexpected Reset	03/06/2022	03/06/2022
							E012207	Twitching	A1107	Data Problem	03/06/2022	03/06/2022
									A160503	Inappropriate or Unexpected Reset	03/06/2022	03/06/2022
							E2403	No Clinical Signs, Symptoms or Conditions	A1107	Data Problem	03/06/2022	03/06/2022
									A160503	Inappropriate or Unexpected Reset	03/06/2022	03/06/2022
							F08	Hospitalization or Prolonged Hospitalization	A1107	Data Problem	03/06/2022	03/06/2022
									A160503	Inappropriate or Unexpected Reset	03/06/2022	03/06/2022
							F1905	Device Revision or Replacement	A1107	Data Problem	03/06/2022	03/06/2022
									A160503	Inappropriate or Unexpected Reset	03/06/2022	03/06/2022
7058	2021	All other reportable incidents	35852	Single-chamber implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E2104	Electric Shock	A070909	Oversensing		
									A071302	Inappropriate/Inadequate Shock/Stimulation		
									A1301	Failure to Read Input Signal		
							E2403	No Clinical Signs, Symptoms or Conditions	A070909	Oversensing		
									A071302	Inappropriate/Inadequate Shock/Stimulation		

									A1301	Failure to Read Input Signal		
							F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A070909	Oversensing		
									A071302	Inappropriate/Inadequate Shock/Stimulation		
									A1301	Failure to Read Input Signal		
7059	2022	All other reportable incidents	35223	Endocardial/interventricular septal pacing lead	CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070102	High Capture Threshold		
									A072201	High Impedance		
							F26	No Health Consequences or Impact	A070102	High Capture Threshold		
									A072201	High Impedance		
7060	2021	All other reportable incidents	47265	Dual-chamber implantable pacemaker, rate-responsive	CT141	Electrodes/Leads and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
					CT145	Cardiovascular devices	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
					CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
					CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
					CT233	Surgical	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
					CT2406	Implantable	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
					CT2407	Active implantable	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
					CT2727	Active implant devices/ancillaries	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
					CT2750	Implantable cardiac electrical stimulators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
					CT2985	Active	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
					CT3001	Implantable metal	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
					CT3002	Implantable synthetic polymer	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
					CT3006	Implantable inorganic material	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
					CT314	Electrically powered	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
					CT321	Long-term surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
					CT449	Generators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
					CT981	Single use	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
					CT983	Surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		

7061	2022	All other reportable incidents	47265	Dual-chamber implantable pacemaker, rate-responsive	CT141	Electrodes/Leads and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	29/09/2022	29/09/2022
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	29/09/2022	29/09/2022
							F1903	Device Explantation	A070504	Premature Discharge of Battery	29/09/2022	29/09/2022
					CT145	Cardiovascular devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	29/09/2022	29/09/2022
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	29/09/2022	29/09/2022
							F1903	Device Explantation	A070504	Premature Discharge of Battery	29/09/2022	29/09/2022
					CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	29/09/2022	29/09/2022
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	29/09/2022	29/09/2022
							F1903	Device Explantation	A070504	Premature Discharge of Battery	29/09/2022	29/09/2022
					CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	29/09/2022	29/09/2022
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	29/09/2022	29/09/2022
							F1903	Device Explantation	A070504	Premature Discharge of Battery	29/09/2022	29/09/2022
					CT233	Surgical	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	29/09/2022	29/09/2022
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	29/09/2022	29/09/2022
							F1903	Device Explantation	A070504	Premature Discharge of Battery	29/09/2022	29/09/2022
					CT2406	Implantable	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	29/09/2022	29/09/2022
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	29/09/2022	29/09/2022
							F1903	Device Explantation	A070504	Premature Discharge of Battery	29/09/2022	29/09/2022
					CT2407	Active implantable	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	29/09/2022	29/09/2022
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	29/09/2022	29/09/2022
							F1903	Device Explantation	A070504	Premature Discharge of Battery	29/09/2022	29/09/2022
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	29/09/2022	29/09/2022
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	29/09/2022	29/09/2022
							F1903	Device Explantation	A070504	Premature Discharge of Battery	29/09/2022	29/09/2022
					CT2727	Active implant devices/ancillaries	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	29/09/2022	29/09/2022
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	29/09/2022	29/09/2022
							F1903	Device Explantation	A070504	Premature Discharge of Battery	29/09/2022	29/09/2022
					CT2750	Implantable cardiac electrical stimulators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	29/09/2022	29/09/2022
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	29/09/2022	29/09/2022
							F1903	Device Explantation	A070504	Premature Discharge of Battery	29/09/2022	29/09/2022
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	29/09/2022	29/09/2022
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	29/09/2022	29/09/2022
							F1903	Device Explantation	A070504	Premature Discharge of Battery	29/09/2022	29/09/2022
					CT2985	Active	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	29/09/2022	29/09/2022
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	29/09/2022	29/09/2022
							F1903	Device Explantation	A070504	Premature Discharge of Battery	29/09/2022	29/09/2022
					CT3001	Implantable metal	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	29/09/2022	29/09/2022
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	29/09/2022	29/09/2022
							F1903	Device Explantation	A070504	Premature Discharge of Battery	29/09/2022	29/09/2022
					CT3002	Implantable synthetic polymer	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	29/09/2022	29/09/2022
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	29/09/2022	29/09/2022
							F1903	Device Explantation	A070504	Premature Discharge of Battery	29/09/2022	29/09/2022
					CT3006	Implantable inorganic material	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	29/09/2022	29/09/2022
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	29/09/2022	29/09/2022
							F1903	Device Explantation	A070504	Premature Discharge of Battery	29/09/2022	29/09/2022
					CT314	Electrically powered	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	29/09/2022	29/09/2022
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	29/09/2022	29/09/2022
							F1903	Device Explantation	A070504	Premature Discharge of Battery	29/09/2022	29/09/2022
					CT321	Long-term surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	29/09/2022	29/09/2022
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	29/09/2022	29/09/2022
							F1903	Device Explantation	A070504	Premature Discharge of Battery	29/09/2022	29/09/2022
					CT449	Generators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	29/09/2022	29/09/2022
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	29/09/2022	29/09/2022
							F1903	Device Explantation	A070504	Premature Discharge of Battery	29/09/2022	29/09/2022
					CT981	Single use	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	29/09/2022	29/09/2022
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	29/09/2022	29/09/2022
							F1903	Device Explantation	A070504	Premature Discharge of Battery	29/09/2022	29/09/2022
					CT983	Surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	29/09/2022	29/09/2022
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	29/09/2022	29/09/2022
							F1903	Device Explantation	A070504	Premature Discharge of Battery	29/09/2022	29/09/2022
7062	2022	All other reportable incidents	36241	Pacemaker/defibrillator lead	CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070101	Failure to Capture		
									A070909	Oversensing		
							F26	No Health Consequences or Impact	A070101	Failure to Capture		
									A070909	Oversensing		
7063	2021	All other reportable incidents	47265	Dual-chamber implantable pacemaker, rate-responsive	CT141	Electrodes/Leads and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT145	Cardiovascular devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT233	Surgical	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT2406	Implantable	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT2407	Active implantable	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT2727	Active implant devices/ancillaries	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT2750	Implantable cardiac electrical stimulators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT2985	Active	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		

					CT3001	Implantable metal	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT3002	Implantable synthetic polymer	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT3006	Implantable inorganic material	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT314	Electrically powered	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT321	Long-term surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT449	Generators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT981	Single use	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT983	Surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
7064	2022	All other reportable incidents	47265	Dual-chamber implantable pacemaker, rate-responsive	CT141	Electrodes/Leads and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A0205	Packaging Problem		
									A070909	Oversensing		
									A090801	Signal Artifact/Noise		
							F26	No Health Consequences or Impact	A0205	Packaging Problem		
									A070909	Oversensing		
									A090801	Signal Artifact/Noise		
					CT145	Cardiovascular devices	E2403	No Clinical Signs, Symptoms or Conditions	A0205	Packaging Problem		
									A070909	Oversensing		
									A090801	Signal Artifact/Noise		
							F26	No Health Consequences or Impact	A0205	Packaging Problem		
									A070909	Oversensing		
									A090801	Signal Artifact/Noise		
					CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A0205	Packaging Problem		
									A070909	Oversensing		
									A090801	Signal Artifact/Noise		
							F26	No Health Consequences or Impact	A0205	Packaging Problem		
									A070909	Oversensing		
									A090801	Signal Artifact/Noise		
					CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A0205	Packaging Problem		
									A070909	Oversensing		
									A090801	Signal Artifact/Noise		
							F26	No Health Consequences or Impact	A0205	Packaging Problem		
									A070909	Oversensing		
									A090801	Signal Artifact/Noise		
					CT233	Surgical	E2403	No Clinical Signs, Symptoms or Conditions	A0205	Packaging Problem		
									A070909	Oversensing		
									A090801	Signal Artifact/Noise		
							F26	No Health Consequences or Impact	A0205	Packaging Problem		
									A070909	Oversensing		
									A090801	Signal Artifact/Noise		
					CT2406	Implantable	E2403	No Clinical Signs, Symptoms or Conditions	A0205	Packaging Problem		
									A070909	Oversensing		
									A090801	Signal Artifact/Noise		
							F26	No Health Consequences or Impact	A0205	Packaging Problem		
									A070909	Oversensing		
									A090801	Signal Artifact/Noise		
					CT2407	Active implantable	E2403	No Clinical Signs, Symptoms or Conditions	A0205	Packaging Problem		
									A070909	Oversensing		
									A090801	Signal Artifact/Noise		
							F26	No Health Consequences or Impact	A0205	Packaging Problem		
									A070909	Oversensing		
									A090801	Signal Artifact/Noise		
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A0205	Packaging Problem		
									A070909	Oversensing		
									A090801	Signal Artifact/Noise		
							F26	No Health Consequences or Impact	A0205	Packaging Problem		
									A070909	Oversensing		
									A090801	Signal Artifact/Noise		
					CT2727	Active implant devices/ancillaries	E2403	No Clinical Signs, Symptoms or Conditions	A0205	Packaging Problem		
									A070909	Oversensing		
									A090801	Signal Artifact/Noise		
							F26	No Health Consequences or Impact	A0205	Packaging Problem		
									A070909	Oversensing		
									A090801	Signal Artifact/Noise		
					CT2750	Implantable cardiac electrical stimulators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A0205	Packaging Problem		
									A070909	Oversensing		
									A090801	Signal Artifact/Noise		
							F26	No Health Consequences or Impact	A0205	Packaging Problem		
									A070909	Oversensing		
									A090801	Signal Artifact/Noise		
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A0205	Packaging Problem		
									A070909	Oversensing		
									A090801	Signal Artifact/Noise		
							F26	No Health Consequences or Impact	A0205	Packaging Problem		
									A070909	Oversensing		
									A090801	Signal Artifact/Noise		
					CT2985	Active	E2403	No Clinical Signs, Symptoms or Conditions	A0205	Packaging Problem		
									A070909	Oversensing		
									A090801	Signal Artifact/Noise		
							F26	No Health Consequences or Impact	A0205	Packaging Problem		
									A070909	Oversensing		
									A090801	Signal Artifact/Noise		

					CT3001	Implantable metal	E2403	No Clinical Signs, Symptoms or Conditions	A0205	Packaging Problem		
									A070909	Oversensing		
									A090801	Signal Artifact/Noise		
							F26	No Health Consequences or Impact	A0205	Packaging Problem		
									A070909	Oversensing		
									A090801	Signal Artifact/Noise		
					CT3002	Implantable synthetic polymer	E2403	No Clinical Signs, Symptoms or Conditions	A0205	Packaging Problem		
									A070909	Oversensing		
									A090801	Signal Artifact/Noise		
							F26	No Health Consequences or Impact	A0205	Packaging Problem		
									A070909	Oversensing		
									A090801	Signal Artifact/Noise		
					CT3006	Implantable inorganic material	E2403	No Clinical Signs, Symptoms or Conditions	A0205	Packaging Problem		
									A070909	Oversensing		
									A090801	Signal Artifact/Noise		
							F26	No Health Consequences or Impact	A0205	Packaging Problem		
									A070909	Oversensing		
									A090801	Signal Artifact/Noise		
					CT314	Electrically powered	E2403	No Clinical Signs, Symptoms or Conditions	A0205	Packaging Problem		
									A070909	Oversensing		
									A090801	Signal Artifact/Noise		
							F26	No Health Consequences or Impact	A0205	Packaging Problem		
									A070909	Oversensing		
									A090801	Signal Artifact/Noise		
					CT321	Long-term surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A0205	Packaging Problem		
									A070909	Oversensing		
									A090801	Signal Artifact/Noise		
							F26	No Health Consequences or Impact	A0205	Packaging Problem		
									A070909	Oversensing		
									A090801	Signal Artifact/Noise		
					CT449	Generators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A0205	Packaging Problem		
									A070909	Oversensing		
									A090801	Signal Artifact/Noise		
							F26	No Health Consequences or Impact	A0205	Packaging Problem		
									A070909	Oversensing		
									A090801	Signal Artifact/Noise		
					CT981	Single use	E2403	No Clinical Signs, Symptoms or Conditions	A0205	Packaging Problem		
									A070909	Oversensing		
									A090801	Signal Artifact/Noise		
							F26	No Health Consequences or Impact	A0205	Packaging Problem		
									A070909	Oversensing		
									A090801	Signal Artifact/Noise		
					CT983	Surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A0205	Packaging Problem		
									A070909	Oversensing		
									A090801	Signal Artifact/Noise		
							F26	No Health Consequences or Impact	A0205	Packaging Problem		
									A070909	Oversensing		
									A090801	Signal Artifact/Noise		
7065	2021	All other reportable incidents	35852	Single-chamber implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery		
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery		
7066	2022	All other reportable incidents	36241	Pacemaker/defibrillator lead	CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A0203	Defective Device	21/06/2022	21/06/2022
									A05	Mechanical Problem	21/06/2022	21/06/2022
									A1502	Positioning Problem	21/06/2022	21/06/2022
							F1908	Prolonged Surgery	A0203	Defective Device	21/06/2022	21/06/2022
									A05	Mechanical Problem	21/06/2022	21/06/2022
									A1502	Positioning Problem	21/06/2022	21/06/2022
7067	2022	All other reportable incidents	47265	Dual-chamber implantable pacemaker, rate-responsive	CT141	Electrodes/Leads and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	14/07/2022	14/07/2022
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	14/07/2022	14/07/2022
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	14/07/2022	14/07/2022
					CT145	Cardiovascular devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	14/07/2022	14/07/2022
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	14/07/2022	14/07/2022
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	14/07/2022	14/07/2022
					CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	14/07/2022	14/07/2022
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	14/07/2022	14/07/2022
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	14/07/2022	14/07/2022
					CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	14/07/2022	14/07/2022
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	14/07/2022	14/07/2022
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	14/07/2022	14/07/2022
					CT233	Surgical	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	14/07/2022	14/07/2022
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	14/07/2022	14/07/2022
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	14/07/2022	14/07/2022
					CT2406	Implantable	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	14/07/2022	14/07/2022
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	14/07/2022	14/07/2022
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	14/07/2022	14/07/2022
					CT2407	Active implantable	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	14/07/2022	14/07/2022
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	14/07/2022	14/07/2022
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	14/07/2022	14/07/2022
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	14/07/2022	14/07/2022
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	14/07/2022	14/07/2022
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	14/07/2022	14/07/2022
					CT2727	Active implant devices/ancillaries	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	14/07/2022	14/07/2022
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	14/07/2022	14/07/2022
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	14/07/2022	14/07/2022
					CT2750	Implantable cardiac electrical stimulators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	14/07/2022	14/07/2022
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	14/07/2022	14/07/2022
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	14/07/2022	14/07/2022
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	14/07/2022	14/07/2022

							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	14/07/2022	14/07/2022
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	14/07/2022	14/07/2022
					CT2985	Active	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	14/07/2022	14/07/2022
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	14/07/2022	14/07/2022
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	14/07/2022	14/07/2022
					CT3001	Implantable metal	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	14/07/2022	14/07/2022
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	14/07/2022	14/07/2022
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	14/07/2022	14/07/2022
					CT3002	Implantable synthetic polymer	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	14/07/2022	14/07/2022
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	14/07/2022	14/07/2022
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	14/07/2022	14/07/2022
					CT3006	Implantable inorganic material	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	14/07/2022	14/07/2022
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	14/07/2022	14/07/2022
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	14/07/2022	14/07/2022
					CT314	Electrically powered	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	14/07/2022	14/07/2022
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	14/07/2022	14/07/2022
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	14/07/2022	14/07/2022
					CT321	Long-term surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	14/07/2022	14/07/2022
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	14/07/2022	14/07/2022
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	14/07/2022	14/07/2022
					CT449	Generators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	14/07/2022	14/07/2022
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	14/07/2022	14/07/2022
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	14/07/2022	14/07/2022
					CT981	Single use	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	14/07/2022	14/07/2022
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	14/07/2022	14/07/2022
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	14/07/2022	14/07/2022
					CT983	Surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	14/07/2022	14/07/2022
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	14/07/2022	14/07/2022
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	14/07/2022	14/07/2022
7068	2021	All other reportable incidents	35223	Endocardial/interventricular septal pacing lead	CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A0401	Break	29/10/2021	29/10/2021
									A070909	Oversensing	29/10/2021	29/10/2021
							F1901	Additional Surgery	A0401	Break	29/10/2021	29/10/2021
									A070909	Oversensing	29/10/2021	29/10/2021
7069	2021	All other reportable incidents	35223	Endocardial/interventricular septal pacing lead	CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A0401	Break	29/10/2021	29/10/2021
									A070909	Oversensing	29/10/2021	29/10/2021
							F1901	Additional Surgery	A0401	Break	29/10/2021	29/10/2021
									A070909	Oversensing	29/10/2021	29/10/2021
7070	2021	All other reportable incidents	47265	Dual-chamber implantable pacemaker, rate-responsive	CT141	Electrodes/Leads and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A0712	Pacing Problem	30/07/2021	30/07/2021
							F1908	Prolonged Surgery	A0712	Pacing Problem	30/07/2021	30/07/2021
					CT145	Cardiovascular devices	E2403	No Clinical Signs, Symptoms or Conditions	A0712	Pacing Problem	30/07/2021	30/07/2021
							F1908	Prolonged Surgery	A0712	Pacing Problem	30/07/2021	30/07/2021
					CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A0712	Pacing Problem	30/07/2021	30/07/2021
							F1908	Prolonged Surgery	A0712	Pacing Problem	30/07/2021	30/07/2021
					CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A0712	Pacing Problem	30/07/2021	30/07/2021
							F1908	Prolonged Surgery	A0712	Pacing Problem	30/07/2021	30/07/2021
					CT233	Surgical	E2403	No Clinical Signs, Symptoms or Conditions	A0712	Pacing Problem	30/07/2021	30/07/2021
							F1908	Prolonged Surgery	A0712	Pacing Problem	30/07/2021	30/07/2021
					CT2406	Implantable	E2403	No Clinical Signs, Symptoms or Conditions	A0712	Pacing Problem	30/07/2021	30/07/2021
							F1908	Prolonged Surgery	A0712	Pacing Problem	30/07/2021	30/07/2021
					CT2407	Active implantable	E2403	No Clinical Signs, Symptoms or Conditions	A0712	Pacing Problem	30/07/2021	30/07/2021
							F1908	Prolonged Surgery	A0712	Pacing Problem	30/07/2021	30/07/2021
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A0712	Pacing Problem	30/07/2021	30/07/2021
							F1908	Prolonged Surgery	A0712	Pacing Problem	30/07/2021	30/07/2021
					CT2727	Active implant devices/ancillaries	E2403	No Clinical Signs, Symptoms or Conditions	A0712	Pacing Problem	30/07/2021	30/07/2021
							F1908	Prolonged Surgery	A0712	Pacing Problem	30/07/2021	30/07/2021
					CT2750	Implantable cardiac electrical stimulators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A0712	Pacing Problem	30/07/2021	30/07/2021
							F1908	Prolonged Surgery	A0712	Pacing Problem	30/07/2021	30/07/2021
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A0712	Pacing Problem	30/07/2021	30/07/2021
							F1908	Prolonged Surgery	A0712	Pacing Problem	30/07/2021	30/07/2021
					CT2985	Active	E2403	No Clinical Signs, Symptoms or Conditions	A0712	Pacing Problem	30/07/2021	30/07/2021
							F1908	Prolonged Surgery	A0712	Pacing Problem	30/07/2021	30/07/2021
					CT3001	Implantable metal	E2403	No Clinical Signs, Symptoms or Conditions	A0712	Pacing Problem	30/07/2021	30/07/2021
							F1908	Prolonged Surgery	A0712	Pacing Problem	30/07/2021	30/07/2021
					CT3002	Implantable synthetic polymer	E2403	No Clinical Signs, Symptoms or Conditions	A0712	Pacing Problem	30/07/2021	30/07/2021
							F1908	Prolonged Surgery	A0712	Pacing Problem	30/07/2021	30/07/2021
					CT3006	Implantable inorganic material	E2403	No Clinical Signs, Symptoms or Conditions	A0712	Pacing Problem	30/07/2021	30/07/2021
							F1908	Prolonged Surgery	A0712	Pacing Problem	30/07/2021	30/07/2021
					CT314	Electrically powered	E2403	No Clinical Signs, Symptoms or Conditions	A0712	Pacing Problem	30/07/2021	30/07/2021
							F1908	Prolonged Surgery	A0712	Pacing Problem	30/07/2021	30/07/2021
					CT321	Long-term surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A0712	Pacing Problem	30/07/2021	30/07/2021
							F1908	Prolonged Surgery	A0712	Pacing Problem	30/07/2021	30/07/2021
					CT449	Generators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A0712	Pacing Problem	30/07/2021	30/07/2021
							F1908	Prolonged Surgery	A0712	Pacing Problem	30/07/2021	30/07/2021
					CT981	Single use	E2403	No Clinical Signs, Symptoms or Conditions	A0712	Pacing Problem	30/07/2021	30/07/2021
							F1908	Prolonged Surgery	A0712	Pacing Problem	30/07/2021	30/07/2021
					CT983	Surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A0712	Pacing Problem	30/07/2021	30/07/2021
							F1908	Prolonged Surgery	A0712	Pacing Problem	30/07/2021	30/07/2021
7071	2021	All other reportable incidents	47270	Cardiac resynchronization therapy implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	10/11/2021	10/11/2021
							F1901	Additional Surgery	A070504	Premature Discharge of Battery	10/11/2021	10/11/2021
7072	2021	All other reportable incidents	35223	Endocardial/interventricular septal pacing lead	CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A0401	Break		
									A070909	Oversensing		
									A090801	Signal Artifact/Noise		
							F26	No Health Consequences or Impact	A0401	Break		
									A070909	Oversensing		
									A090801	Signal Artifact/Noise		
7073	2022	All other reportable incidents	47270	Cardiac resynchronization therapy implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	27/06/2022	27/06/2022
							F1901	Additional Surgery	A070504	Premature Discharge of Battery	27/06/2022	27/06/2022
7074	2021	All other reportable incidents	36241	Pacemaker/defibrillator lead	CT1684	Cardiac pacemakers and associated devices	E0604	Cardiac Perforation	A24	No Apparent Device Problem		

							E0605	Cardiac Tamponade	A24	No Apparent Device Problem		
							E2321	Low Blood Pressure/ Hypotension	A24	No Apparent Device Problem		
							E233001	Chest Pain	A24	No Apparent Device Problem		
							E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem		
							F1203	Life Threatening Illness or Injury	A24	No Apparent Device Problem		
							F15	Recognized Device or Procedural Complication	A24	No Apparent Device Problem		
							F2203	Imaging Required	A24	No Apparent Device Problem		
							F23	Unexpected Medical Intervention	A24	No Apparent Device Problem		
7075	2021	All other reportable incidents	35223	Endocardial/interventricular septal pacing lead	CT1684	Cardiac pacemakers and associated devices	E0112	Dizziness	A040101	Fracture		
									A070101	Failure to Capture		
							E2403	No Clinical Signs, Symptoms or Conditions	A040101	Fracture		
									A070101	Failure to Capture		
							F08	Hospitalization or Prolonged Hospitalization	A040101	Fracture		
									A070101	Failure to Capture		
							F11	Non-Serious Injury/ Illness/ Impairment	A040101	Fracture		
									A070101	Failure to Capture		
							F1905	Device Revision or Replacement	A040101	Fracture		
									A070101	Failure to Capture		
7076	2021	All other reportable incidents	37265	Dual-chamber implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E060110	Ventricular Fibrillation	A0713	Defibrillation/Stimulation Problem		
							E160304	Skull Fracture	A0713	Defibrillation/Stimulation Problem		
							E2403	No Clinical Signs, Symptoms or Conditions	A0713	Defibrillation/Stimulation Problem		
							F08	Hospitalization or Prolonged Hospitalization	A0713	Defibrillation/Stimulation Problem		
							F1203	Life Threatening Illness or Injury	A0713	Defibrillation/Stimulation Problem		
							F23	Unexpected Medical Intervention	A0713	Defibrillation/Stimulation Problem		
7077	2022	All other reportable incidents	17882	Physiologic-monitoring defibrillation system	CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A090206	No Display/Image		
							F27	Problem Identified During Non-Clinical Procedure	A090206	No Display/Image		
7078	2022	All other reportable incidents	17882	Physiologic-monitoring defibrillation system	CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A090206	No Display/Image		
							F27	Problem Identified During Non-Clinical Procedure	A090206	No Display/Image		
7079	2022	All other reportable incidents	17882	Physiologic-monitoring defibrillation system	CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A090206	No Display/Image		
							F27	Problem Identified During Non-Clinical Procedure	A090206	No Display/Image		
7080	2022	All other reportable incidents	17882	Physiologic-monitoring defibrillation system	CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A090206	No Display/Image		
							F27	Problem Identified During Non-Clinical Procedure	A090206	No Display/Image		
7081	2021	All other reportable incidents	35852	Single-chamber implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem	08/11/2021	08/11/2021
									A070504	Premature Discharge of Battery	08/11/2021	08/11/2021
							F08	Hospitalization or Prolonged Hospitalization	A0705	Battery Problem	08/11/2021	08/11/2021
									A070504	Premature Discharge of Battery	08/11/2021	08/11/2021
							F12	Serious Injury/ Illness/ Impairment	A0705	Battery Problem	08/11/2021	08/11/2021
									A070504	Premature Discharge of Battery	08/11/2021	08/11/2021
							F1905	Device Revision or Replacement	A0705	Battery Problem	08/11/2021	08/11/2021
									A070504	Premature Discharge of Battery	08/11/2021	08/11/2021
7082	2022	All other reportable incidents	47265	Dual-chamber implantable pacemaker, rate-responsive	CT141	Electrodes/Leads and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem	15/09/2022	15/09/2022
									A070504	Premature Discharge of Battery	15/09/2022	15/09/2022
							F08	Hospitalization or Prolonged Hospitalization	A0705	Battery Problem	15/09/2022	15/09/2022
									A070504	Premature Discharge of Battery	15/09/2022	15/09/2022
							F1905	Device Revision or Replacement	A0705	Battery Problem	15/09/2022	15/09/2022
									A070504	Premature Discharge of Battery	15/09/2022	15/09/2022
					CT145	Cardiovascular devices	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem	15/09/2022	15/09/2022
									A070504	Premature Discharge of Battery	15/09/2022	15/09/2022
							F08	Hospitalization or Prolonged Hospitalization	A0705	Battery Problem	15/09/2022	15/09/2022
									A070504	Premature Discharge of Battery	15/09/2022	15/09/2022
							F1905	Device Revision or Replacement	A0705	Battery Problem	15/09/2022	15/09/2022
									A070504	Premature Discharge of Battery	15/09/2022	15/09/2022
					CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem	15/09/2022	15/09/2022
									A070504	Premature Discharge of Battery	15/09/2022	15/09/2022
							F08	Hospitalization or Prolonged Hospitalization	A0705	Battery Problem	15/09/2022	15/09/2022
									A070504	Premature Discharge of Battery	15/09/2022	15/09/2022
							F1905	Device Revision or Replacement	A0705	Battery Problem	15/09/2022	15/09/2022
									A070504	Premature Discharge of Battery	15/09/2022	15/09/2022
					CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem	15/09/2022	15/09/2022
									A070504	Premature Discharge of Battery	15/09/2022	15/09/2022
							F08	Hospitalization or Prolonged Hospitalization	A0705	Battery Problem	15/09/2022	15/09/2022
									A070504	Premature Discharge of Battery	15/09/2022	15/09/2022
							F1905	Device Revision or Replacement	A0705	Battery Problem	15/09/2022	15/09/2022
									A070504	Premature Discharge of Battery	15/09/2022	15/09/2022
					CT233	Surgical	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem	15/09/2022	15/09/2022
									A070504	Premature Discharge of Battery	15/09/2022	15/09/2022
							F08	Hospitalization or Prolonged Hospitalization	A0705	Battery Problem	15/09/2022	15/09/2022
									A070504	Premature Discharge of Battery	15/09/2022	15/09/2022
							F1905	Device Revision or Replacement	A0705	Battery Problem	15/09/2022	15/09/2022
									A070504	Premature Discharge of Battery	15/09/2022	15/09/2022
					CT2406	Implantable	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem	15/09/2022	15/09/2022
									A070504	Premature Discharge of Battery	15/09/2022	15/09/2022
							F08	Hospitalization or Prolonged Hospitalization	A0705	Battery Problem	15/09/2022	15/09/2022
									A070504	Premature Discharge of Battery	15/09/2022	15/09/2022
							F1905	Device Revision or Replacement	A0705	Battery Problem	15/09/2022	15/09/2022
									A070504	Premature Discharge of Battery	15/09/2022	15/09/2022
					CT2407	Active implantable	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem	15/09/2022	15/09/2022
									A070504	Premature Discharge of Battery	15/09/2022	15/09/2022
							F08	Hospitalization or Prolonged Hospitalization	A0705	Battery Problem	15/09/2022	15/09/2022
									A070504	Premature Discharge of Battery	15/09/2022	15/09/2022
							F1905	Device Revision or Replacement	A0705	Battery Problem	15/09/2022	15/09/2022
									A070504	Premature Discharge of Battery	15/09/2022	15/09/2022
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem	15/09/2022	15/09/2022
									A070504	Premature Discharge of Battery	15/09/2022	15/09/2022
							F08	Hospitalization or Prolonged Hospitalization	A0705	Battery Problem	15/09/2022	15/09/2022
									A070504	Premature Discharge of Battery	15/09/2022	15/09/2022
							F1905	Device Revision or Replacement	A0705	Battery Problem	15/09/2022	15/09/2022
									A070504	Premature Discharge of Battery	15/09/2022	15/09/2022

					CT2727	Active implant devices/ancillaries	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem	15/09/2022	15/09/2022
									A070504	Premature Discharge of Battery	15/09/2022	15/09/2022
							F08	Hospitalization or Prolonged Hospitalization	A0705	Battery Problem	15/09/2022	15/09/2022
									A070504	Premature Discharge of Battery	15/09/2022	15/09/2022
							F1905	Device Revision or Replacement	A0705	Battery Problem	15/09/2022	15/09/2022
									A070504	Premature Discharge of Battery	15/09/2022	15/09/2022
					CT2750	Implantable cardiac electrical stimulators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem	15/09/2022	15/09/2022
									A070504	Premature Discharge of Battery	15/09/2022	15/09/2022
							F08	Hospitalization or Prolonged Hospitalization	A0705	Battery Problem	15/09/2022	15/09/2022
									A070504	Premature Discharge of Battery	15/09/2022	15/09/2022
							F1905	Device Revision or Replacement	A0705	Battery Problem	15/09/2022	15/09/2022
									A070504	Premature Discharge of Battery	15/09/2022	15/09/2022
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem	15/09/2022	15/09/2022
									A070504	Premature Discharge of Battery	15/09/2022	15/09/2022
							F08	Hospitalization or Prolonged Hospitalization	A0705	Battery Problem	15/09/2022	15/09/2022
									A070504	Premature Discharge of Battery	15/09/2022	15/09/2022
							F1905	Device Revision or Replacement	A0705	Battery Problem	15/09/2022	15/09/2022
									A070504	Premature Discharge of Battery	15/09/2022	15/09/2022
					CT2985	Active	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem	15/09/2022	15/09/2022
									A070504	Premature Discharge of Battery	15/09/2022	15/09/2022
							F08	Hospitalization or Prolonged Hospitalization	A0705	Battery Problem	15/09/2022	15/09/2022
									A070504	Premature Discharge of Battery	15/09/2022	15/09/2022
							F1905	Device Revision or Replacement	A0705	Battery Problem	15/09/2022	15/09/2022
									A070504	Premature Discharge of Battery	15/09/2022	15/09/2022
					CT3001	Implantable metal	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem	15/09/2022	15/09/2022
									A070504	Premature Discharge of Battery	15/09/2022	15/09/2022
							F08	Hospitalization or Prolonged Hospitalization	A0705	Battery Problem	15/09/2022	15/09/2022
									A070504	Premature Discharge of Battery	15/09/2022	15/09/2022
							F1905	Device Revision or Replacement	A0705	Battery Problem	15/09/2022	15/09/2022
									A070504	Premature Discharge of Battery	15/09/2022	15/09/2022
					CT3002	Implantable synthetic polymer	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem	15/09/2022	15/09/2022
									A070504	Premature Discharge of Battery	15/09/2022	15/09/2022
							F08	Hospitalization or Prolonged Hospitalization	A0705	Battery Problem	15/09/2022	15/09/2022
									A070504	Premature Discharge of Battery	15/09/2022	15/09/2022
							F1905	Device Revision or Replacement	A0705	Battery Problem	15/09/2022	15/09/2022
									A070504	Premature Discharge of Battery	15/09/2022	15/09/2022
					CT3006	Implantable inorganic material	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem	15/09/2022	15/09/2022
									A070504	Premature Discharge of Battery	15/09/2022	15/09/2022
							F08	Hospitalization or Prolonged Hospitalization	A0705	Battery Problem	15/09/2022	15/09/2022
									A070504	Premature Discharge of Battery	15/09/2022	15/09/2022
							F1905	Device Revision or Replacement	A0705	Battery Problem	15/09/2022	15/09/2022
									A070504	Premature Discharge of Battery	15/09/2022	15/09/2022
					CT314	Electrically powered	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem	15/09/2022	15/09/2022
									A070504	Premature Discharge of Battery	15/09/2022	15/09/2022
							F08	Hospitalization or Prolonged Hospitalization	A0705	Battery Problem	15/09/2022	15/09/2022
									A070504	Premature Discharge of Battery	15/09/2022	15/09/2022
							F1905	Device Revision or Replacement	A0705	Battery Problem	15/09/2022	15/09/2022
									A070504	Premature Discharge of Battery	15/09/2022	15/09/2022
					CT321	Long-term surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem	15/09/2022	15/09/2022
									A070504	Premature Discharge of Battery	15/09/2022	15/09/2022
							F08	Hospitalization or Prolonged Hospitalization	A0705	Battery Problem	15/09/2022	15/09/2022
									A070504	Premature Discharge of Battery	15/09/2022	15/09/2022
							F1905	Device Revision or Replacement	A0705	Battery Problem	15/09/2022	15/09/2022
									A070504	Premature Discharge of Battery	15/09/2022	15/09/2022
					CT449	Generators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem	15/09/2022	15/09/2022
									A070504	Premature Discharge of Battery	15/09/2022	15/09/2022
							F08	Hospitalization or Prolonged Hospitalization	A0705	Battery Problem	15/09/2022	15/09/2022
									A070504	Premature Discharge of Battery	15/09/2022	15/09/2022
							F1905	Device Revision or Replacement	A0705	Battery Problem	15/09/2022	15/09/2022
									A070504	Premature Discharge of Battery	15/09/2022	15/09/2022
					CT981	Single use	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem	15/09/2022	15/09/2022
									A070504	Premature Discharge of Battery	15/09/2022	15/09/2022
							F08	Hospitalization or Prolonged Hospitalization	A0705	Battery Problem	15/09/2022	15/09/2022
									A070504	Premature Discharge of Battery	15/09/2022	15/09/2022
							F1905	Device Revision or Replacement	A0705	Battery Problem	15/09/2022	15/09/2022
									A070504	Premature Discharge of Battery	15/09/2022	15/09/2022
					CT983	Surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem	15/09/2022	15/09/2022
									A070504	Premature Discharge of Battery	15/09/2022	15/09/2022
							F08	Hospitalization or Prolonged Hospitalization	A0705	Battery Problem	15/09/2022	15/09/2022
									A070504	Premature Discharge of Battery	15/09/2022	15/09/2022
							F1905	Device Revision or Replacement	A0705	Battery Problem	15/09/2022	15/09/2022
									A070504	Premature Discharge of Battery	15/09/2022	15/09/2022
7083	2022	All other reportable incidents	47267	Single-chamber implantable pacemaker, rate-responsive	CT141	Electrodes/Leads and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	01/07/2022	01/07/2022
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery	01/07/2022	01/07/2022
					CT145	Cardiovascular devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	01/07/2022	01/07/2022
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery	01/07/2022	01/07/2022
					CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	01/07/2022	01/07/2022
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery	01/07/2022	01/07/2022
					CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	01/07/2022	01/07/2022
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery	01/07/2022	01/07/2022
					CT233	Surgical	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	01/07/2022	01/07/2022
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery	01/07/2022	01/07/2022
					CT2406	Implantable	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	01/07/2022	01/07/2022
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery	01/07/2022	01/07/2022
					CT2407	Active implantable	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	01/07/2022	01/07/2022
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery	01/07/2022	01/07/2022
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	01/07/2022	01/07/2022
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery	01/07/2022	01/07/2022

					CT2727	Active implant devices/ancillaries	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	01/07/2022	01/07/2022
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery	01/07/2022	01/07/2022
					CT2750	Implantable cardiac electrical stimulators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	01/07/2022	01/07/2022
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery	01/07/2022	01/07/2022
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	01/07/2022	01/07/2022
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery	01/07/2022	01/07/2022
					CT2985	Active	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	01/07/2022	01/07/2022
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery	01/07/2022	01/07/2022
					CT3001	Implantable metal	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	01/07/2022	01/07/2022
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery	01/07/2022	01/07/2022
					CT3002	Implantable synthetic polymer	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	01/07/2022	01/07/2022
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery	01/07/2022	01/07/2022
					CT3006	Implantable inorganic material	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	01/07/2022	01/07/2022
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery	01/07/2022	01/07/2022
					CT314	Electrically powered	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	01/07/2022	01/07/2022
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery	01/07/2022	01/07/2022
					CT321	Long-term surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	01/07/2022	01/07/2022
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery	01/07/2022	01/07/2022
					CT449	Generators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	01/07/2022	01/07/2022
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery	01/07/2022	01/07/2022
					CT981	Single use	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	01/07/2022	01/07/2022
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery	01/07/2022	01/07/2022
					CT983	Surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	01/07/2022	01/07/2022
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery	01/07/2022	01/07/2022
7084	2022	All other reportable incidents	47265	Dual-chamber implantable pacemaker, rate-responsive	CT141	Electrodes/Leads and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem	13/07/2022	13/07/2022
									A070504	Premature Discharge of Battery	13/07/2022	13/07/2022
							F08	Hospitalization or Prolonged Hospitalization	A0705	Battery Problem	13/07/2022	13/07/2022
									A070504	Premature Discharge of Battery	13/07/2022	13/07/2022
							F1905	Device Revision or Replacement	A0705	Battery Problem	13/07/2022	13/07/2022
									A070504	Premature Discharge of Battery	13/07/2022	13/07/2022
					CT145	Cardiovascular devices	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem	13/07/2022	13/07/2022
									A070504	Premature Discharge of Battery	13/07/2022	13/07/2022
							F08	Hospitalization or Prolonged Hospitalization	A0705	Battery Problem	13/07/2022	13/07/2022
									A070504	Premature Discharge of Battery	13/07/2022	13/07/2022
							F1905	Device Revision or Replacement	A0705	Battery Problem	13/07/2022	13/07/2022
									A070504	Premature Discharge of Battery	13/07/2022	13/07/2022
					CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem	13/07/2022	13/07/2022
									A070504	Premature Discharge of Battery	13/07/2022	13/07/2022
							F08	Hospitalization or Prolonged Hospitalization	A0705	Battery Problem	13/07/2022	13/07/2022
									A070504	Premature Discharge of Battery	13/07/2022	13/07/2022
							F1905	Device Revision or Replacement	A0705	Battery Problem	13/07/2022	13/07/2022
									A070504	Premature Discharge of Battery	13/07/2022	13/07/2022
					CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem	13/07/2022	13/07/2022
									A070504	Premature Discharge of Battery	13/07/2022	13/07/2022
							F08	Hospitalization or Prolonged Hospitalization	A0705	Battery Problem	13/07/2022	13/07/2022
									A070504	Premature Discharge of Battery	13/07/2022	13/07/2022
							F1905	Device Revision or Replacement	A0705	Battery Problem	13/07/2022	13/07/2022
									A070504	Premature Discharge of Battery	13/07/2022	13/07/2022
					CT233	Surgical	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem	13/07/2022	13/07/2022
									A070504	Premature Discharge of Battery	13/07/2022	13/07/2022
							F08	Hospitalization or Prolonged Hospitalization	A0705	Battery Problem	13/07/2022	13/07/2022
									A070504	Premature Discharge of Battery	13/07/2022	13/07/2022
							F1905	Device Revision or Replacement	A0705	Battery Problem	13/07/2022	13/07/2022
									A070504	Premature Discharge of Battery	13/07/2022	13/07/2022
					CT2406	Implantable	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem	13/07/2022	13/07/2022
									A070504	Premature Discharge of Battery	13/07/2022	13/07/2022
							F08	Hospitalization or Prolonged Hospitalization	A0705	Battery Problem	13/07/2022	13/07/2022
									A070504	Premature Discharge of Battery	13/07/2022	13/07/2022
							F1905	Device Revision or Replacement	A0705	Battery Problem	13/07/2022	13/07/2022
									A070504	Premature Discharge of Battery	13/07/2022	13/07/2022
					CT2407	Active implantable	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem	13/07/2022	13/07/2022
									A070504	Premature Discharge of Battery	13/07/2022	13/07/2022
							F08	Hospitalization or Prolonged Hospitalization	A0705	Battery Problem	13/07/2022	13/07/2022
									A070504	Premature Discharge of Battery	13/07/2022	13/07/2022
							F1905	Device Revision or Replacement	A0705	Battery Problem	13/07/2022	13/07/2022
									A070504	Premature Discharge of Battery	13/07/2022	13/07/2022
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem	13/07/2022	13/07/2022
									A070504	Premature Discharge of Battery	13/07/2022	13/07/2022
							F08	Hospitalization or Prolonged Hospitalization	A0705	Battery Problem	13/07/2022	13/07/2022
									A070504	Premature Discharge of Battery	13/07/2022	13/07/2022
							F1905	Device Revision or Replacement	A0705	Battery Problem	13/07/2022	13/07/2022
									A070504	Premature Discharge of Battery	13/07/2022	13/07/2022
					CT2727	Active implant devices/ancillaries	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem	13/07/2022	13/07/2022
									A070504	Premature Discharge of Battery	13/07/2022	13/07/2022
							F08	Hospitalization or Prolonged Hospitalization	A0705	Battery Problem	13/07/2022	13/07/2022
									A070504	Premature Discharge of Battery	13/07/2022	13/07/2022
							F1905	Device Revision or Replacement	A0705	Battery Problem	13/07/2022	13/07/2022
									A070504	Premature Discharge of Battery	13/07/2022	13/07/2022
					CT2750	Implantable cardiac electrical stimulators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem	13/07/2022	13/07/2022
									A070504	Premature Discharge of Battery	13/07/2022	13/07/2022
							F08	Hospitalization or Prolonged Hospitalization	A0705	Battery Problem	13/07/2022	13/07/2022
									A070504	Premature Discharge of Battery	13/07/2022	13/07/2022
							F1905	Device Revision or Replacement	A0705	Battery Problem	13/07/2022	13/07/2022
									A070504	Premature Discharge of Battery	13/07/2022	13/07/2022
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem	13/07/2022	13/07/2022
									A070504	Premature Discharge of Battery	13/07/2022	13/07/2022
							F08	Hospitalization or Prolonged Hospitalization	A0705	Battery Problem	13/07/2022	13/07/2022
									A070504	Premature Discharge of Battery	13/07/2022	13/07/2022

							F1905	Device Revision or Replacement	A0705	Battery Problem	13/07/2022	13/07/2022
									A070504	Premature Discharge of Battery	13/07/2022	13/07/2022
					CT2985	Active	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem	13/07/2022	13/07/2022
									A070504	Premature Discharge of Battery	13/07/2022	13/07/2022
							F08	Hospitalization or Prolonged Hospitalization	A0705	Battery Problem	13/07/2022	13/07/2022
									A070504	Premature Discharge of Battery	13/07/2022	13/07/2022
							F1905	Device Revision or Replacement	A0705	Battery Problem	13/07/2022	13/07/2022
									A070504	Premature Discharge of Battery	13/07/2022	13/07/2022
					CT3001	Implantable metal	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem	13/07/2022	13/07/2022
									A070504	Premature Discharge of Battery	13/07/2022	13/07/2022
							F08	Hospitalization or Prolonged Hospitalization	A0705	Battery Problem	13/07/2022	13/07/2022
									A070504	Premature Discharge of Battery	13/07/2022	13/07/2022
							F1905	Device Revision or Replacement	A0705	Battery Problem	13/07/2022	13/07/2022
									A070504	Premature Discharge of Battery	13/07/2022	13/07/2022
					CT3002	Implantable synthetic polymer	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem	13/07/2022	13/07/2022
									A070504	Premature Discharge of Battery	13/07/2022	13/07/2022
							F08	Hospitalization or Prolonged Hospitalization	A0705	Battery Problem	13/07/2022	13/07/2022
									A070504	Premature Discharge of Battery	13/07/2022	13/07/2022
							F1905	Device Revision or Replacement	A0705	Battery Problem	13/07/2022	13/07/2022
									A070504	Premature Discharge of Battery	13/07/2022	13/07/2022
					CT3006	Implantable inorganic material	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem	13/07/2022	13/07/2022
									A070504	Premature Discharge of Battery	13/07/2022	13/07/2022
							F08	Hospitalization or Prolonged Hospitalization	A0705	Battery Problem	13/07/2022	13/07/2022
									A070504	Premature Discharge of Battery	13/07/2022	13/07/2022
							F1905	Device Revision or Replacement	A0705	Battery Problem	13/07/2022	13/07/2022
									A070504	Premature Discharge of Battery	13/07/2022	13/07/2022
					CT314	Electrically powered	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem	13/07/2022	13/07/2022
									A070504	Premature Discharge of Battery	13/07/2022	13/07/2022
							F08	Hospitalization or Prolonged Hospitalization	A0705	Battery Problem	13/07/2022	13/07/2022
									A070504	Premature Discharge of Battery	13/07/2022	13/07/2022
							F1905	Device Revision or Replacement	A0705	Battery Problem	13/07/2022	13/07/2022
									A070504	Premature Discharge of Battery	13/07/2022	13/07/2022
					CT321	Long-term surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem	13/07/2022	13/07/2022
									A070504	Premature Discharge of Battery	13/07/2022	13/07/2022
							F08	Hospitalization or Prolonged Hospitalization	A0705	Battery Problem	13/07/2022	13/07/2022
									A070504	Premature Discharge of Battery	13/07/2022	13/07/2022
							F1905	Device Revision or Replacement	A0705	Battery Problem	13/07/2022	13/07/2022
									A070504	Premature Discharge of Battery	13/07/2022	13/07/2022
					CT449	Generators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem	13/07/2022	13/07/2022
									A070504	Premature Discharge of Battery	13/07/2022	13/07/2022
							F08	Hospitalization or Prolonged Hospitalization	A0705	Battery Problem	13/07/2022	13/07/2022
									A070504	Premature Discharge of Battery	13/07/2022	13/07/2022
							F1905	Device Revision or Replacement	A0705	Battery Problem	13/07/2022	13/07/2022
									A070504	Premature Discharge of Battery	13/07/2022	13/07/2022
					CT981	Single use	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem	13/07/2022	13/07/2022
									A070504	Premature Discharge of Battery	13/07/2022	13/07/2022
							F08	Hospitalization or Prolonged Hospitalization	A0705	Battery Problem	13/07/2022	13/07/2022
									A070504	Premature Discharge of Battery	13/07/2022	13/07/2022
							F1905	Device Revision or Replacement	A0705	Battery Problem	13/07/2022	13/07/2022
									A070504	Premature Discharge of Battery	13/07/2022	13/07/2022
					CT983	Surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem	13/07/2022	13/07/2022
									A070504	Premature Discharge of Battery	13/07/2022	13/07/2022
							F08	Hospitalization or Prolonged Hospitalization	A0705	Battery Problem	13/07/2022	13/07/2022
									A070504	Premature Discharge of Battery	13/07/2022	13/07/2022
							F1905	Device Revision or Replacement	A0705	Battery Problem	13/07/2022	13/07/2022
									A070504	Premature Discharge of Battery	13/07/2022	13/07/2022
7085	2022	All other reportable incidents	47270	Cardiac resynchronization therapy implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	24/06/2022	24/06/2022
							F1901	Additional Surgery	A070504	Premature Discharge of Battery	24/06/2022	24/06/2022
7086	2021	All other reportable incidents	47265	Dual-chamber implantable pacemaker, rate-responsive	CT141	Electrodes/Leads and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
							F24	Insufficient Health Impact Information	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
					CT145	Cardiovascular devices	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
							F24	Insufficient Health Impact Information	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
					CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
							F24	Insufficient Health Impact Information	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
					CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
							F24	Insufficient Health Impact Information	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
					CT233	Surgical	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
							F24	Insufficient Health Impact Information	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
					CT2406	Implantable	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
							F24	Insufficient Health Impact Information	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
					CT2407	Active implantable	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
							F24	Insufficient Health Impact Information	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		

							F24	Insufficient Health Impact Information	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
					CT2727	Active implant devices/ancillaries	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
							F24	Insufficient Health Impact Information	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
					CT2750	Implantable cardiac electrical stimulators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
							F24	Insufficient Health Impact Information	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
							F24	Insufficient Health Impact Information	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
					CT2985	Active	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
							F24	Insufficient Health Impact Information	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
					CT3001	Implantable metal	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
							F24	Insufficient Health Impact Information	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
					CT3002	Implantable synthetic polymer	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
							F24	Insufficient Health Impact Information	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
					CT3006	Implantable inorganic material	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
							F24	Insufficient Health Impact Information	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
					CT314	Electrically powered	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
							F24	Insufficient Health Impact Information	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
					CT321	Long-term surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
							F24	Insufficient Health Impact Information	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
					CT449	Generators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
							F24	Insufficient Health Impact Information	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
					CT981	Single use	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
							F24	Insufficient Health Impact Information	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
					CT983	Surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
							F24	Insufficient Health Impact Information	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
7087	2021	All other reportable incidents	47265	Dual-chamber implantable pacemaker, rate-responsive	CT141	Electrodes/Leads and associated devices	E011903	Syncope/Fainting	A0712	Pacing Problem	04/10/2021	04/10/2021
									A0908	Incorrect, Inadequate or Imprecise Result or Readings	04/10/2021	04/10/2021
							E2403	No Clinical Signs, Symptoms or Conditions	A0712	Pacing Problem	04/10/2021	04/10/2021
									A0908	Incorrect, Inadequate or Imprecise Result or Readings	04/10/2021	04/10/2021
							F08	Hospitalization or Prolonged Hospitalization	A0712	Pacing Problem	04/10/2021	04/10/2021
									A0908	Incorrect, Inadequate or Imprecise Result or Readings	04/10/2021	04/10/2021
							F1905	Device Revision or Replacement	A0712	Pacing Problem	04/10/2021	04/10/2021
									A0908	Incorrect, Inadequate or Imprecise Result or Readings	04/10/2021	04/10/2021
					CT145	Cardiovascular devices	E011903	Syncope/Fainting	A0712	Pacing Problem	04/10/2021	04/10/2021
									A0908	Incorrect, Inadequate or Imprecise Result or Readings	04/10/2021	04/10/2021
							E2403	No Clinical Signs, Symptoms or Conditions	A0712	Pacing Problem	04/10/2021	04/10/2021
									A0908	Incorrect, Inadequate or Imprecise Result or Readings	04/10/2021	04/10/2021
							F08	Hospitalization or Prolonged Hospitalization	A0712	Pacing Problem	04/10/2021	04/10/2021
									A0908	Incorrect, Inadequate or Imprecise Result or Readings	04/10/2021	04/10/2021
							F1905	Device Revision or Replacement	A0712	Pacing Problem	04/10/2021	04/10/2021
									A0908	Incorrect, Inadequate or Imprecise Result or Readings	04/10/2021	04/10/2021
					CT1684	Cardiac pacemakers and associated devices	E011903	Syncope/Fainting	A0712	Pacing Problem	04/10/2021	04/10/2021
									A0908	Incorrect, Inadequate or Imprecise Result or Readings	04/10/2021	04/10/2021
							E2403	No Clinical Signs, Symptoms or Conditions	A0712	Pacing Problem	04/10/2021	04/10/2021
									A0908	Incorrect, Inadequate or Imprecise Result or Readings	04/10/2021	04/10/2021
							F08	Hospitalization or Prolonged Hospitalization	A0712	Pacing Problem	04/10/2021	04/10/2021
									A0908	Incorrect, Inadequate or Imprecise Result or Readings	04/10/2021	04/10/2021
							F1905	Device Revision or Replacement	A0712	Pacing Problem	04/10/2021	04/10/2021
									A0908	Incorrect, Inadequate or Imprecise Result or Readings	04/10/2021	04/10/2021
					CT1685	Cardiac pacemakers	E011903	Syncope/Fainting	A0712	Pacing Problem	04/10/2021	04/10/2021
									A0908	Incorrect, Inadequate or Imprecise Result or Readings	04/10/2021	04/10/2021
							E2403	No Clinical Signs, Symptoms or Conditions	A0712	Pacing Problem	04/10/2021	04/10/2021

									A0908	Incorrect, Inadequate or Imprecise Result or Readings	04/10/2021	04/10/2021
							F08	Hospitalization or Prolonged Hospitalization	A0712	Pacing Problem	04/10/2021	04/10/2021
									A0908	Incorrect, Inadequate or Imprecise Result or Readings	04/10/2021	04/10/2021
							F1905	Device Revision or Replacement	A0712	Pacing Problem	04/10/2021	04/10/2021
									A0908	Incorrect, Inadequate or Imprecise Result or Readings	04/10/2021	04/10/2021
					CT233	Surgical	E011903	Syncope/Fainting	A0712	Pacing Problem	04/10/2021	04/10/2021
									A0908	Incorrect, Inadequate or Imprecise Result or Readings	04/10/2021	04/10/2021
							E2403	No Clinical Signs, Symptoms or Conditions	A0712	Pacing Problem	04/10/2021	04/10/2021
									A0908	Incorrect, Inadequate or Imprecise Result or Readings	04/10/2021	04/10/2021
							F08	Hospitalization or Prolonged Hospitalization	A0712	Pacing Problem	04/10/2021	04/10/2021
									A0908	Incorrect, Inadequate or Imprecise Result or Readings	04/10/2021	04/10/2021
							F1905	Device Revision or Replacement	A0712	Pacing Problem	04/10/2021	04/10/2021
									A0908	Incorrect, Inadequate or Imprecise Result or Readings	04/10/2021	04/10/2021
					CT2406	Implantable	E011903	Syncope/Fainting	A0712	Pacing Problem	04/10/2021	04/10/2021
									A0908	Incorrect, Inadequate or Imprecise Result or Readings	04/10/2021	04/10/2021
							E2403	No Clinical Signs, Symptoms or Conditions	A0712	Pacing Problem	04/10/2021	04/10/2021
									A0908	Incorrect, Inadequate or Imprecise Result or Readings	04/10/2021	04/10/2021
							F08	Hospitalization or Prolonged Hospitalization	A0712	Pacing Problem	04/10/2021	04/10/2021
									A0908	Incorrect, Inadequate or Imprecise Result or Readings	04/10/2021	04/10/2021
							F1905	Device Revision or Replacement	A0712	Pacing Problem	04/10/2021	04/10/2021
									A0908	Incorrect, Inadequate or Imprecise Result or Readings	04/10/2021	04/10/2021
					CT2407	Active implantable	E011903	Syncope/Fainting	A0712	Pacing Problem	04/10/2021	04/10/2021
									A0908	Incorrect, Inadequate or Imprecise Result or Readings	04/10/2021	04/10/2021
							E2403	No Clinical Signs, Symptoms or Conditions	A0712	Pacing Problem	04/10/2021	04/10/2021
									A0908	Incorrect, Inadequate or Imprecise Result or Readings	04/10/2021	04/10/2021
							F08	Hospitalization or Prolonged Hospitalization	A0712	Pacing Problem	04/10/2021	04/10/2021
									A0908	Incorrect, Inadequate or Imprecise Result or Readings	04/10/2021	04/10/2021
							F1905	Device Revision or Replacement	A0712	Pacing Problem	04/10/2021	04/10/2021
									A0908	Incorrect, Inadequate or Imprecise Result or Readings	04/10/2021	04/10/2021
					CT257	Implantable cardiac pacemakers	E011903	Syncope/Fainting	A0712	Pacing Problem	04/10/2021	04/10/2021
									A0908	Incorrect, Inadequate or Imprecise Result or Readings	04/10/2021	04/10/2021
							E2403	No Clinical Signs, Symptoms or Conditions	A0712	Pacing Problem	04/10/2021	04/10/2021
									A0908	Incorrect, Inadequate or Imprecise Result or Readings	04/10/2021	04/10/2021
							F08	Hospitalization or Prolonged Hospitalization	A0712	Pacing Problem	04/10/2021	04/10/2021
									A0908	Incorrect, Inadequate or Imprecise Result or Readings	04/10/2021	04/10/2021
							F1905	Device Revision or Replacement	A0712	Pacing Problem	04/10/2021	04/10/2021
									A0908	Incorrect, Inadequate or Imprecise Result or Readings	04/10/2021	04/10/2021
					CT2727	Active implant devices/ancillaries	E011903	Syncope/Fainting	A0712	Pacing Problem	04/10/2021	04/10/2021
									A0908	Incorrect, Inadequate or Imprecise Result or Readings	04/10/2021	04/10/2021
							E2403	No Clinical Signs, Symptoms or Conditions	A0712	Pacing Problem	04/10/2021	04/10/2021
									A0908	Incorrect, Inadequate or Imprecise Result or Readings	04/10/2021	04/10/2021
							F08	Hospitalization or Prolonged Hospitalization	A0712	Pacing Problem	04/10/2021	04/10/2021
									A0908	Incorrect, Inadequate or Imprecise Result or Readings	04/10/2021	04/10/2021
							F1905	Device Revision or Replacement	A0712	Pacing Problem	04/10/2021	04/10/2021
									A0908	Incorrect, Inadequate or Imprecise Result or Readings	04/10/2021	04/10/2021
					CT2750	Implantable cardiac electrical stimulators and associated devices	E011903	Syncope/Fainting	A0712	Pacing Problem	04/10/2021	04/10/2021
									A0908	Incorrect, Inadequate or Imprecise Result or Readings	04/10/2021	04/10/2021
							E2403	No Clinical Signs, Symptoms or Conditions	A0712	Pacing Problem	04/10/2021	04/10/2021
									A0908	Incorrect, Inadequate or Imprecise Result or Readings	04/10/2021	04/10/2021
							F08	Hospitalization or Prolonged Hospitalization	A0712	Pacing Problem	04/10/2021	04/10/2021
									A0908	Incorrect, Inadequate or Imprecise Result or Readings	04/10/2021	04/10/2021
							F1905	Device Revision or Replacement	A0712	Pacing Problem	04/10/2021	04/10/2021
									A0908	Incorrect, Inadequate or Imprecise Result or Readings	04/10/2021	04/10/2021
					CT2866	Implantable cardiac electrical stimulators	E011903	Syncope/Fainting	A0712	Pacing Problem	04/10/2021	04/10/2021
									A0908	Incorrect, Inadequate or Imprecise Result or Readings	04/10/2021	04/10/2021
							E2403	No Clinical Signs, Symptoms or Conditions	A0712	Pacing Problem	04/10/2021	04/10/2021
									A0908	Incorrect, Inadequate or Imprecise Result or Readings	04/10/2021	04/10/2021
							F08	Hospitalization or Prolonged Hospitalization	A0712	Pacing Problem	04/10/2021	04/10/2021
									A0908	Incorrect, Inadequate or Imprecise Result or Readings	04/10/2021	04/10/2021
							F1905	Device Revision or Replacement	A0712	Pacing Problem	04/10/2021	04/10/2021
									A0908	Incorrect, Inadequate or Imprecise Result or Readings	04/10/2021	04/10/2021
					CT2985	Active	E011903	Syncope/Fainting	A0712	Pacing Problem	04/10/2021	04/10/2021
									A0908	Incorrect, Inadequate or Imprecise Result or Readings	04/10/2021	04/10/2021

							E2403	No Clinical Signs, Symptoms or Conditions	A0712	Pacing Problem	04/10/2021	04/10/2021
									A0908	Incorrect, Inadequate or Imprecise Result or Readings	04/10/2021	04/10/2021
							F08	Hospitalization or Prolonged Hospitalization	A0712	Pacing Problem	04/10/2021	04/10/2021
									A0908	Incorrect, Inadequate or Imprecise Result or Readings	04/10/2021	04/10/2021
							F1905	Device Revision or Replacement	A0712	Pacing Problem	04/10/2021	04/10/2021
									A0908	Incorrect, Inadequate or Imprecise Result or Readings	04/10/2021	04/10/2021
					CT3001	Implantable metal	E011903	Syncope/Fainting	A0712	Pacing Problem	04/10/2021	04/10/2021
									A0908	Incorrect, Inadequate or Imprecise Result or Readings	04/10/2021	04/10/2021
							E2403	No Clinical Signs, Symptoms or Conditions	A0712	Pacing Problem	04/10/2021	04/10/2021
									A0908	Incorrect, Inadequate or Imprecise Result or Readings	04/10/2021	04/10/2021
							F08	Hospitalization or Prolonged Hospitalization	A0712	Pacing Problem	04/10/2021	04/10/2021
									A0908	Incorrect, Inadequate or Imprecise Result or Readings	04/10/2021	04/10/2021
							F1905	Device Revision or Replacement	A0712	Pacing Problem	04/10/2021	04/10/2021
									A0908	Incorrect, Inadequate or Imprecise Result or Readings	04/10/2021	04/10/2021
					CT3002	Implantable synthetic polymer	E011903	Syncope/Fainting	A0712	Pacing Problem	04/10/2021	04/10/2021
									A0908	Incorrect, Inadequate or Imprecise Result or Readings	04/10/2021	04/10/2021
							E2403	No Clinical Signs, Symptoms or Conditions	A0712	Pacing Problem	04/10/2021	04/10/2021
									A0908	Incorrect, Inadequate or Imprecise Result or Readings	04/10/2021	04/10/2021
							F08	Hospitalization or Prolonged Hospitalization	A0712	Pacing Problem	04/10/2021	04/10/2021
									A0908	Incorrect, Inadequate or Imprecise Result or Readings	04/10/2021	04/10/2021
							F1905	Device Revision or Replacement	A0712	Pacing Problem	04/10/2021	04/10/2021
									A0908	Incorrect, Inadequate or Imprecise Result or Readings	04/10/2021	04/10/2021
					CT3006	Implantable inorganic material	E011903	Syncope/Fainting	A0712	Pacing Problem	04/10/2021	04/10/2021
									A0908	Incorrect, Inadequate or Imprecise Result or Readings	04/10/2021	04/10/2021
							E2403	No Clinical Signs, Symptoms or Conditions	A0712	Pacing Problem	04/10/2021	04/10/2021
									A0908	Incorrect, Inadequate or Imprecise Result or Readings	04/10/2021	04/10/2021
							F08	Hospitalization or Prolonged Hospitalization	A0712	Pacing Problem	04/10/2021	04/10/2021
									A0908	Incorrect, Inadequate or Imprecise Result or Readings	04/10/2021	04/10/2021
							F1905	Device Revision or Replacement	A0712	Pacing Problem	04/10/2021	04/10/2021
									A0908	Incorrect, Inadequate or Imprecise Result or Readings	04/10/2021	04/10/2021
					CT314	Electrically powered	E011903	Syncope/Fainting	A0712	Pacing Problem	04/10/2021	04/10/2021
									A0908	Incorrect, Inadequate or Imprecise Result or Readings	04/10/2021	04/10/2021
							E2403	No Clinical Signs, Symptoms or Conditions	A0712	Pacing Problem	04/10/2021	04/10/2021
									A0908	Incorrect, Inadequate or Imprecise Result or Readings	04/10/2021	04/10/2021
							F08	Hospitalization or Prolonged Hospitalization	A0712	Pacing Problem	04/10/2021	04/10/2021
									A0908	Incorrect, Inadequate or Imprecise Result or Readings	04/10/2021	04/10/2021
							F1905	Device Revision or Replacement	A0712	Pacing Problem	04/10/2021	04/10/2021
									A0908	Incorrect, Inadequate or Imprecise Result or Readings	04/10/2021	04/10/2021
					CT321	Long-term surgical invasive	E011903	Syncope/Fainting	A0712	Pacing Problem	04/10/2021	04/10/2021
									A0908	Incorrect, Inadequate or Imprecise Result or Readings	04/10/2021	04/10/2021
							E2403	No Clinical Signs, Symptoms or Conditions	A0712	Pacing Problem	04/10/2021	04/10/2021
									A0908	Incorrect, Inadequate or Imprecise Result or Readings	04/10/2021	04/10/2021
							F08	Hospitalization or Prolonged Hospitalization	A0712	Pacing Problem	04/10/2021	04/10/2021
									A0908	Incorrect, Inadequate or Imprecise Result or Readings	04/10/2021	04/10/2021
							F1905	Device Revision or Replacement	A0712	Pacing Problem	04/10/2021	04/10/2021
									A0908	Incorrect, Inadequate or Imprecise Result or Readings	04/10/2021	04/10/2021
					CT449	Generators and associated devices	E011903	Syncope/Fainting	A0712	Pacing Problem	04/10/2021	04/10/2021
									A0908	Incorrect, Inadequate or Imprecise Result or Readings	04/10/2021	04/10/2021
							E2403	No Clinical Signs, Symptoms or Conditions	A0712	Pacing Problem	04/10/2021	04/10/2021
									A0908	Incorrect, Inadequate or Imprecise Result or Readings	04/10/2021	04/10/2021
							F08	Hospitalization or Prolonged Hospitalization	A0712	Pacing Problem	04/10/2021	04/10/2021
									A0908	Incorrect, Inadequate or Imprecise Result or Readings	04/10/2021	04/10/2021
							F1905	Device Revision or Replacement	A0712	Pacing Problem	04/10/2021	04/10/2021
									A0908	Incorrect, Inadequate or Imprecise Result or Readings	04/10/2021	04/10/2021
					CT981	Single use	E011903	Syncope/Fainting	A0712	Pacing Problem	04/10/2021	04/10/2021
									A0908	Incorrect, Inadequate or Imprecise Result or Readings	04/10/2021	04/10/2021
							E2403	No Clinical Signs, Symptoms or Conditions	A0712	Pacing Problem	04/10/2021	04/10/2021
									A0908	Incorrect, Inadequate or Imprecise Result or Readings	04/10/2021	04/10/2021
							F08	Hospitalization or Prolonged Hospitalization	A0712	Pacing Problem	04/10/2021	04/10/2021
									A0908	Incorrect, Inadequate or Imprecise Result or Readings	04/10/2021	04/10/2021
							F1905	Device Revision or Replacement	A0712	Pacing Problem	04/10/2021	04/10/2021
									A0908	Incorrect, Inadequate or Imprecise Result or Readings	04/10/2021	04/10/2021
					CT983	Surgical invasive	E011903	Syncope/Fainting	A0712	Pacing Problem	04/10/2021	04/10/2021

								A0908	Incorrect, Inadequate or Imprecise Result or Readings	04/10/2021	04/10/2021
						E2403	No Clinical Signs, Symptoms or Conditions	A0712	Pacing Problem	04/10/2021	04/10/2021
								A0908	Incorrect, Inadequate or Imprecise Result or Readings	04/10/2021	04/10/2021
						F08	Hospitalization or Prolonged Hospitalization	A0712	Pacing Problem	04/10/2021	04/10/2021
								A0908	Incorrect, Inadequate or Imprecise Result or Readings	04/10/2021	04/10/2021
						F1905	Device Revision or Replacement	A0712	Pacing Problem	04/10/2021	04/10/2021
								A0908	Incorrect, Inadequate or Imprecise Result or Readings	04/10/2021	04/10/2021
7088	2021	All other reportable incidents	47265	Dual-chamber implantable pacemaker, rate-responsive	CT141	Electrodes/Leads and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem	14/10/2021	14/10/2021
							F08	Hospitalization or Prolonged Hospitalization	A0705	Battery Problem	14/10/2021	14/10/2021
							F1905	Device Revision or Replacement	A0705	Battery Problem	14/10/2021	14/10/2021
					CT145	Cardiovascular devices	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem	14/10/2021	14/10/2021
							F08	Hospitalization or Prolonged Hospitalization	A0705	Battery Problem	14/10/2021	14/10/2021
							F1905	Device Revision or Replacement	A0705	Battery Problem	14/10/2021	14/10/2021
					CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem	14/10/2021	14/10/2021
							F08	Hospitalization or Prolonged Hospitalization	A0705	Battery Problem	14/10/2021	14/10/2021
							F1905	Device Revision or Replacement	A0705	Battery Problem	14/10/2021	14/10/2021
					CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem	14/10/2021	14/10/2021
							F08	Hospitalization or Prolonged Hospitalization	A0705	Battery Problem	14/10/2021	14/10/2021
							F1905	Device Revision or Replacement	A0705	Battery Problem	14/10/2021	14/10/2021
					CT233	Surgical	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem	14/10/2021	14/10/2021
							F08	Hospitalization or Prolonged Hospitalization	A0705	Battery Problem	14/10/2021	14/10/2021
							F1905	Device Revision or Replacement	A0705	Battery Problem	14/10/2021	14/10/2021
					CT2406	Implantable	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem	14/10/2021	14/10/2021
							F08	Hospitalization or Prolonged Hospitalization	A0705	Battery Problem	14/10/2021	14/10/2021
							F1905	Device Revision or Replacement	A0705	Battery Problem	14/10/2021	14/10/2021
					CT2407	Active implantable	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem	14/10/2021	14/10/2021
							F08	Hospitalization or Prolonged Hospitalization	A0705	Battery Problem	14/10/2021	14/10/2021
							F1905	Device Revision or Replacement	A0705	Battery Problem	14/10/2021	14/10/2021
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem	14/10/2021	14/10/2021
							F08	Hospitalization or Prolonged Hospitalization	A0705	Battery Problem	14/10/2021	14/10/2021
							F1905	Device Revision or Replacement	A0705	Battery Problem	14/10/2021	14/10/2021
					CT2727	Active implant devices/ancillaries	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem	14/10/2021	14/10/2021
							F08	Hospitalization or Prolonged Hospitalization	A0705	Battery Problem	14/10/2021	14/10/2021
							F1905	Device Revision or Replacement	A0705	Battery Problem	14/10/2021	14/10/2021
					CT2750	Implantable cardiac electrical stimulators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem	14/10/2021	14/10/2021
							F08	Hospitalization or Prolonged Hospitalization	A0705	Battery Problem	14/10/2021	14/10/2021
							F1905	Device Revision or Replacement	A0705	Battery Problem	14/10/2021	14/10/2021
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem	14/10/2021	14/10/2021
							F08	Hospitalization or Prolonged Hospitalization	A0705	Battery Problem	14/10/2021	14/10/2021
							F1905	Device Revision or Replacement	A0705	Battery Problem	14/10/2021	14/10/2021
					CT2985	Active	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem	14/10/2021	14/10/2021
							F08	Hospitalization or Prolonged Hospitalization	A0705	Battery Problem	14/10/2021	14/10/2021
							F1905	Device Revision or Replacement	A0705	Battery Problem	14/10/2021	14/10/2021
					CT3001	Implantable metal	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem	14/10/2021	14/10/2021
							F08	Hospitalization or Prolonged Hospitalization	A0705	Battery Problem	14/10/2021	14/10/2021
							F1905	Device Revision or Replacement	A0705	Battery Problem	14/10/2021	14/10/2021
					CT3002	Implantable synthetic polymer	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem	14/10/2021	14/10/2021
							F08	Hospitalization or Prolonged Hospitalization	A0705	Battery Problem	14/10/2021	14/10/2021
							F1905	Device Revision or Replacement	A0705	Battery Problem	14/10/2021	14/10/2021
					CT3006	Implantable inorganic material	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem	14/10/2021	14/10/2021
							F08	Hospitalization or Prolonged Hospitalization	A0705	Battery Problem	14/10/2021	14/10/2021
							F1905	Device Revision or Replacement	A0705	Battery Problem	14/10/2021	14/10/2021
					CT314	Electrically powered	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem	14/10/2021	14/10/2021
							F08	Hospitalization or Prolonged Hospitalization	A0705	Battery Problem	14/10/2021	14/10/2021
							F1905	Device Revision or Replacement	A0705	Battery Problem	14/10/2021	14/10/2021
					CT321	Long-term surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem	14/10/2021	14/10/2021
							F08	Hospitalization or Prolonged Hospitalization	A0705	Battery Problem	14/10/2021	14/10/2021
							F1905	Device Revision or Replacement	A0705	Battery Problem	14/10/2021	14/10/2021
					CT449	Generators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem	14/10/2021	14/10/2021
							F08	Hospitalization or Prolonged Hospitalization	A0705	Battery Problem	14/10/2021	14/10/2021
							F1905	Device Revision or Replacement	A0705	Battery Problem	14/10/2021	14/10/2021
					CT981	Single use	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem	14/10/2021	14/10/2021
							F08	Hospitalization or Prolonged Hospitalization	A0705	Battery Problem	14/10/2021	14/10/2021
							F1905	Device Revision or Replacement	A0705	Battery Problem	14/10/2021	14/10/2021
					CT983	Surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem	14/10/2021	14/10/2021
							F08	Hospitalization or Prolonged Hospitalization	A0705	Battery Problem	14/10/2021	14/10/2021
							F1905	Device Revision or Replacement	A0705	Battery Problem	14/10/2021	14/10/2021
7089	2021	All other reportable incidents	47265	Dual-chamber implantable pacemaker, rate-responsive	CT141	Electrodes/Leads and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	29/08/2022	29/08/2022
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	29/08/2022	29/08/2022
							F1903	Device Explantation	A070504	Premature Discharge of Battery	29/08/2022	29/08/2022
					CT145	Cardiovascular devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	29/08/2022	29/08/2022
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	29/08/2022	29/08/2022
							F1903	Device Explantation	A070504	Premature Discharge of Battery	29/08/2022	29/08/2022
					CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	29/08/2022	29/08/2022
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	29/08/2022	29/08/2022
							F1903	Device Explantation	A070504	Premature Discharge of Battery	29/08/2022	29/08/2022
					CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	29/08/2022	29/08/2022
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	29/08/2022	29/08/2022
							F1903	Device Explantation	A070504	Premature Discharge of Battery	29/08/2022	29/08/2022
					CT233	Surgical	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	29/08/2022	29/08/2022
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	29/08/2022	29/08/2022
							F1903	Device Explantation	A070504	Premature Discharge of Battery	29/08/2022	29/08/2022
					CT2406	Implantable	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	29/08/2022	29/08/2022
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	29/08/2022	29/08/2022
							F1903	Device Explantation	A070504	Premature Discharge of Battery	29/08/2022	29/08/2022

					CT2407	Active implantable	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	29/08/2022	29/08/2022
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	29/08/2022	29/08/2022
							F1903	Device Explantation	A070504	Premature Discharge of Battery	29/08/2022	29/08/2022
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	29/08/2022	29/08/2022
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	29/08/2022	29/08/2022
							F1903	Device Explantation	A070504	Premature Discharge of Battery	29/08/2022	29/08/2022
					CT2727	Active implant devices/ancillaries	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	29/08/2022	29/08/2022
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	29/08/2022	29/08/2022
							F1903	Device Explantation	A070504	Premature Discharge of Battery	29/08/2022	29/08/2022
					CT2750	Implantable cardiac electrical stimulators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	29/08/2022	29/08/2022
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	29/08/2022	29/08/2022
							F1903	Device Explantation	A070504	Premature Discharge of Battery	29/08/2022	29/08/2022
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	29/08/2022	29/08/2022
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	29/08/2022	29/08/2022
							F1903	Device Explantation	A070504	Premature Discharge of Battery	29/08/2022	29/08/2022
					CT2985	Active	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	29/08/2022	29/08/2022
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	29/08/2022	29/08/2022
							F1903	Device Explantation	A070504	Premature Discharge of Battery	29/08/2022	29/08/2022
					CT3001	Implantable metal	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	29/08/2022	29/08/2022
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	29/08/2022	29/08/2022
							F1903	Device Explantation	A070504	Premature Discharge of Battery	29/08/2022	29/08/2022
					CT3002	Implantable synthetic polymer	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	29/08/2022	29/08/2022
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	29/08/2022	29/08/2022
							F1903	Device Explantation	A070504	Premature Discharge of Battery	29/08/2022	29/08/2022
					CT3006	Implantable inorganic material	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	29/08/2022	29/08/2022
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	29/08/2022	29/08/2022
							F1903	Device Explantation	A070504	Premature Discharge of Battery	29/08/2022	29/08/2022
					CT314	Electrically powered	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	29/08/2022	29/08/2022
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	29/08/2022	29/08/2022
							F1903	Device Explantation	A070504	Premature Discharge of Battery	29/08/2022	29/08/2022
					CT321	Long-term surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	29/08/2022	29/08/2022
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	29/08/2022	29/08/2022
							F1903	Device Explantation	A070504	Premature Discharge of Battery	29/08/2022	29/08/2022
					CT449	Generators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	29/08/2022	29/08/2022
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	29/08/2022	29/08/2022
							F1903	Device Explantation	A070504	Premature Discharge of Battery	29/08/2022	29/08/2022
					CT981	Single use	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	29/08/2022	29/08/2022
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	29/08/2022	29/08/2022
							F1903	Device Explantation	A070504	Premature Discharge of Battery	29/08/2022	29/08/2022
					CT983	Surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	29/08/2022	29/08/2022
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	29/08/2022	29/08/2022
							F1903	Device Explantation	A070504	Premature Discharge of Battery	29/08/2022	29/08/2022
7090	2022	All other reportable incidents	17882	Physiologic-monitoring defibrillation system	CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A0511	Structural Problem		
							F27	Problem Identified During Non-Clinical Procedure	A0511	Structural Problem		
7091	2022	All other reportable incidents	47265	Dual-chamber implantable pacemaker, rate-responsive	CT141	Electrodes/Leads and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070909	Oversensing		
									A072201	High Impedance		
							F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A070909	Oversensing		
									A072201	High Impedance		
					CT145	Cardiovascular devices	E2403	No Clinical Signs, Symptoms or Conditions	A070909	Oversensing		
									A072201	High Impedance		
							F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A070909	Oversensing		
									A072201	High Impedance		
					CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070909	Oversensing		
									A072201	High Impedance		
							F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A070909	Oversensing		
									A072201	High Impedance		
					CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A070909	Oversensing		
									A072201	High Impedance		
							F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A070909	Oversensing		
									A072201	High Impedance		
					CT233	Surgical	E2403	No Clinical Signs, Symptoms or Conditions	A070909	Oversensing		
									A072201	High Impedance		
							F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A070909	Oversensing		
									A072201	High Impedance		
					CT2406	Implantable	E2403	No Clinical Signs, Symptoms or Conditions	A070909	Oversensing		
									A072201	High Impedance		
							F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A070909	Oversensing		
									A072201	High Impedance		
					CT2407	Active implantable	E2403	No Clinical Signs, Symptoms or Conditions	A070909	Oversensing		
									A072201	High Impedance		
							F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A070909	Oversensing		
									A072201	High Impedance		
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A070909	Oversensing		
									A072201	High Impedance		
							F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A070909	Oversensing		
									A072201	High Impedance		
					CT2727	Active implant devices/ancillaries	E2403	No Clinical Signs, Symptoms or Conditions	A070909	Oversensing		
									A072201	High Impedance		
							F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A070909	Oversensing		
									A072201	High Impedance		
					CT2750	Implantable cardiac electrical stimulators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070909	Oversensing		
									A072201	High Impedance		
							F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A070909	Oversensing		
									A072201	High Impedance		
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A070909	Oversensing		
									A072201	High Impedance		
							F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A070909	Oversensing		
									A072201	High Impedance		

					CT2985	Active	E2403	No Clinical Signs, Symptoms or Conditions	A070909	Oversensing		
									A072201	High Impedance		
							F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A070909	Oversensing		
									A072201	High Impedance		
					CT3001	Implantable metal	E2403	No Clinical Signs, Symptoms or Conditions	A070909	Oversensing		
									A072201	High Impedance		
							F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A070909	Oversensing		
									A072201	High Impedance		
					CT3002	Implantable synthetic polymer	E2403	No Clinical Signs, Symptoms or Conditions	A070909	Oversensing		
									A072201	High Impedance		
							F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A070909	Oversensing		
									A072201	High Impedance		
					CT3006	Implantable inorganic material	E2403	No Clinical Signs, Symptoms or Conditions	A070909	Oversensing		
									A072201	High Impedance		
							F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A070909	Oversensing		
									A072201	High Impedance		
					CT314	Electrically powered	E2403	No Clinical Signs, Symptoms or Conditions	A070909	Oversensing		
									A072201	High Impedance		
							F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A070909	Oversensing		
									A072201	High Impedance		
					CT321	Long-term surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A070909	Oversensing		
									A072201	High Impedance		
							F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A070909	Oversensing		
									A072201	High Impedance		
					CT449	Generators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070909	Oversensing		
									A072201	High Impedance		
							F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A070909	Oversensing		
									A072201	High Impedance		
					CT981	Single use	E2403	No Clinical Signs, Symptoms or Conditions	A070909	Oversensing		
									A072201	High Impedance		
							F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A070909	Oversensing		
									A072201	High Impedance		
					CT983	Surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A070909	Oversensing		
									A072201	High Impedance		
							F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A070909	Oversensing		
									A072201	High Impedance		
7092	2022	All other reportable incidents	35852	Single-chamber implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem	29/09/2022	29/09/2022
									A070504	Premature Discharge of Battery	29/09/2022	29/09/2022
							F08	Hospitalization or Prolonged Hospitalization	A0705	Battery Problem	29/09/2022	29/09/2022
									A070504	Premature Discharge of Battery	29/09/2022	29/09/2022
							F1905	Device Revision or Replacement	A0705	Battery Problem	29/09/2022	29/09/2022
									A070504	Premature Discharge of Battery	29/09/2022	29/09/2022
7093	2021	All other reportable incidents	17882	Physiologic-monitoring defibrillation system	CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A0719	Unexpected Shutdown		
							F27	Problem Identified During Non-Clinical Procedure	A0719	Unexpected Shutdown		
7094	2022	All other reportable incidents	47265	Dual-chamber implantable pacemaker, rate-responsive	CT141	Electrodes/Leads and associated devices	E011903	Syncope/Fainting	A071102	Failure to Interrogate	27/06/2022	27/06/2022
									A071204	No Pacing	27/06/2022	27/06/2022
							E0601	Arrhythmia	A071102	Failure to Interrogate	27/06/2022	27/06/2022
									A071204	No Pacing	27/06/2022	27/06/2022
							E0717	Dyspnea	A071102	Failure to Interrogate	27/06/2022	27/06/2022
									A071204	No Pacing	27/06/2022	27/06/2022
							E2403	No Clinical Signs, Symptoms or Conditions	A071102	Failure to Interrogate	27/06/2022	27/06/2022
									A071204	No Pacing	27/06/2022	27/06/2022
							F1901	Additional Surgery	A071102	Failure to Interrogate	27/06/2022	27/06/2022
									A071204	No Pacing	27/06/2022	27/06/2022
					CT145	Cardiovascular devices	E011903	Syncope/Fainting	A071102	Failure to Interrogate	27/06/2022	27/06/2022
									A071204	No Pacing	27/06/2022	27/06/2022
							E0601	Arrhythmia	A071102	Failure to Interrogate	27/06/2022	27/06/2022
									A071204	No Pacing	27/06/2022	27/06/2022
							E0717	Dyspnea	A071102	Failure to Interrogate	27/06/2022	27/06/2022
									A071204	No Pacing	27/06/2022	27/06/2022
							E2403	No Clinical Signs, Symptoms or Conditions	A071102	Failure to Interrogate	27/06/2022	27/06/2022
									A071204	No Pacing	27/06/2022	27/06/2022
							F1901	Additional Surgery	A071102	Failure to Interrogate	27/06/2022	27/06/2022
									A071204	No Pacing	27/06/2022	27/06/2022
					CT1684	Cardiac pacemakers and associated devices	E011903	Syncope/Fainting	A071102	Failure to Interrogate	27/06/2022	27/06/2022
									A071204	No Pacing	27/06/2022	27/06/2022
							E0601	Arrhythmia	A071102	Failure to Interrogate	27/06/2022	27/06/2022
									A071204	No Pacing	27/06/2022	27/06/2022
							E0717	Dyspnea	A071102	Failure to Interrogate	27/06/2022	27/06/2022
									A071204	No Pacing	27/06/2022	27/06/2022
							E2403	No Clinical Signs, Symptoms or Conditions	A071102	Failure to Interrogate	27/06/2022	27/06/2022
									A071204	No Pacing	27/06/2022	27/06/2022
							F1901	Additional Surgery	A071102	Failure to Interrogate	27/06/2022	27/06/2022
									A071204	No Pacing	27/06/2022	27/06/2022
					CT1685	Cardiac pacemakers	E011903	Syncope/Fainting	A071102	Failure to Interrogate	27/06/2022	27/06/2022
									A071204	No Pacing	27/06/2022	27/06/2022
							E0601	Arrhythmia	A071102	Failure to Interrogate	27/06/2022	27/06/2022
									A071204	No Pacing	27/06/2022	27/06/2022
							E0717	Dyspnea	A071102	Failure to Interrogate	27/06/2022	27/06/2022
									A071204	No Pacing	27/06/2022	27/06/2022
							E2403	No Clinical Signs, Symptoms or Conditions	A071102	Failure to Interrogate	27/06/2022	27/06/2022
									A071204	No Pacing	27/06/2022	27/06/2022
							F1901	Additional Surgery	A071102	Failure to Interrogate	27/06/2022	27/06/2022
									A071204	No Pacing	27/06/2022	27/06/2022
					CT233	Surgical	E011903	Syncope/Fainting	A071102	Failure to Interrogate	27/06/2022	27/06/2022
									A071204	No Pacing	27/06/2022	27/06/2022
							E0601	Arrhythmia	A071102	Failure to Interrogate	27/06/2022	27/06/2022
									A071204	No Pacing	27/06/2022	27/06/2022

							E0717	Dyspnea	A071102	Failure to Interrogate	27/06/2022	27/06/2022
									A071204	No Pacing	27/06/2022	27/06/2022
							E2403	No Clinical Signs, Symptoms or Conditions	A071102	Failure to Interrogate	27/06/2022	27/06/2022
									A071204	No Pacing	27/06/2022	27/06/2022
							F1901	Additional Surgery	A071102	Failure to Interrogate	27/06/2022	27/06/2022
									A071204	No Pacing	27/06/2022	27/06/2022
						CT2406	Implantable	E011903	Syncope/Fainting	A071102	Failure to Interrogate	27/06/2022	27/06/2022
										A071204	No Pacing	27/06/2022	27/06/2022
								E0601	Arrhythmia	A071102	Failure to Interrogate	27/06/2022	27/06/2022
										A071204	No Pacing	27/06/2022	27/06/2022
								E0717	Dyspnea	A071102	Failure to Interrogate	27/06/2022	27/06/2022
										A071204	No Pacing	27/06/2022	27/06/2022
								E2403	No Clinical Signs, Symptoms or Conditions	A071102	Failure to Interrogate	27/06/2022	27/06/2022
										A071204	No Pacing	27/06/2022	27/06/2022
								F1901	Additional Surgery	A071102	Failure to Interrogate	27/06/2022	27/06/2022
										A071204	No Pacing	27/06/2022	27/06/2022
						CT2407	Active implantable	E011903	Syncope/Fainting	A071102	Failure to Interrogate	27/06/2022	27/06/2022
										A071204	No Pacing	27/06/2022	27/06/2022
								E0601	Arrhythmia	A071102	Failure to Interrogate	27/06/2022	27/06/2022
										A071204	No Pacing	27/06/2022	27/06/2022
								E0717	Dyspnea	A071102	Failure to Interrogate	27/06/2022	27/06/2022
										A071204	No Pacing	27/06/2022	27/06/2022
								E2403	No Clinical Signs, Symptoms or Conditions	A071102	Failure to Interrogate	27/06/2022	27/06/2022
										A071204	No Pacing	27/06/2022	27/06/2022
								F1901	Additional Surgery	A071102	Failure to Interrogate	27/06/2022	27/06/2022
										A071204	No Pacing	27/06/2022	27/06/2022
						CT257	Implantable cardiac pacemakers	E011903	Syncope/Fainting	A071102	Failure to Interrogate	27/06/2022	27/06/2022
										A071204	No Pacing	27/06/2022	27/06/2022
								E0601	Arrhythmia	A071102	Failure to Interrogate	27/06/2022	27/06/2022
										A071204	No Pacing	27/06/2022	27/06/2022
								E0717	Dyspnea	A071102	Failure to Interrogate	27/06/2022	27/06/2022
										A071204	No Pacing	27/06/2022	27/06/2022
								E2403	No Clinical Signs, Symptoms or Conditions	A071102	Failure to Interrogate	27/06/2022	27/06/2022
										A071204	No Pacing	27/06/2022	27/06/2022
								F1901	Additional Surgery	A071102	Failure to Interrogate	27/06/2022	27/06/2022
										A071204	No Pacing	27/06/2022	27/06/2022
						CT2727	Active implant devices/ancillaries	E011903	Syncope/Fainting	A071102	Failure to Interrogate	27/06/2022	27/06/2022
										A071204	No Pacing	27/06/2022	27/06/2022
								E0601	Arrhythmia	A071102	Failure to Interrogate	27/06/2022	27/06/2022
										A071204	No Pacing	27/06/2022	27/06/2022
								E0717	Dyspnea	A071102	Failure to Interrogate	27/06/2022	27/06/2022
										A071204	No Pacing	27/06/2022	27/06/2022
								E2403	No Clinical Signs, Symptoms or Conditions	A071102	Failure to Interrogate	27/06/2022	27/06/2022
										A071204	No Pacing	27/06/2022	27/06/2022
								F1901	Additional Surgery	A071102	Failure to Interrogate	27/06/2022	27/06/2022
										A071204	No Pacing	27/06/2022	27/06/2022
						CT2750	Implantable cardiac electrical stimulators and associated devices	E011903	Syncope/Fainting	A071102	Failure to Interrogate	27/06/2022	27/06/2022
										A071204	No Pacing	27/06/2022	27/06/2022
								E0601	Arrhythmia	A071102	Failure to Interrogate	27/06/2022	27/06/2022
										A071204	No Pacing	27/06/2022	27/06/2022
								E0717	Dyspnea	A071102	Failure to Interrogate	27/06/2022	27/06/2022
										A071204	No Pacing	27/06/2022	27/06/2022
								E2403	No Clinical Signs, Symptoms or Conditions	A071102	Failure to Interrogate	27/06/2022	27/06/2022
										A071204	No Pacing	27/06/2022	27/06/2022
								F1901	Additional Surgery	A071102	Failure to Interrogate	27/06/2022	27/06/2022
										A071204	No Pacing	27/06/2022	27/06/2022
						CT2866	Implantable cardiac electrical stimulators	E011903	Syncope/Fainting	A071102	Failure to Interrogate	27/06/2022	27/06/2022
										A071204	No Pacing	27/06/2022	27/06/2022
								E0601	Arrhythmia	A071102	Failure to Interrogate	27/06/2022	27/06/2022
										A071204	No Pacing	27/06/2022	27/06/2022
								E0717	Dyspnea	A071102	Failure to Interrogate	27/06/2022	27/06/2022
										A071204	No Pacing	27/06/2022	27/06/2022
								E2403	No Clinical Signs, Symptoms or Conditions	A071102	Failure to Interrogate	27/06/2022	27/06/2022
										A071204	No Pacing	27/06/2022	27/06/2022
								F1901	Additional Surgery	A071102	Failure to Interrogate	27/06/2022	27/06/2022
										A071204	No Pacing	27/06/2022	27/06/2022
						CT2985	Active	E011903	Syncope/Fainting	A071102	Failure to Interrogate	27/06/2022	27/06/2022
										A071204	No Pacing	27/06/2022	27/06/2022
								E0601	Arrhythmia	A071102	Failure to Interrogate	27/06/2022	27/06/2022
										A071204	No Pacing	27/06/2022	27/06/2022
								E0717	Dyspnea	A071102	Failure to Interrogate	27/06/2022	27/06/2022
										A071204	No Pacing	27/06/2022	27/06/2022
								E2403	No Clinical Signs, Symptoms or Conditions	A071102	Failure to Interrogate	27/06/2022	27/06/2022
										A071204	No Pacing	27/06/2022	27/06/2022
								F1901	Additional Surgery	A071102	Failure to Interrogate	27/06/2022	27/06/2022
										A071204	No Pacing	27/06/2022	27/06/2022
						CT3001	Implantable metal	E011903	Syncope/Fainting	A071102	Failure to Interrogate	27/06/2022	27/06/2022
										A071204	No Pacing	27/06/2022	27/06/2022
								E0601	Arrhythmia	A071102	Failure to Interrogate	27/06/2022	27/06/2022
										A071204	No Pacing	27/06/2022	27/06/2022
								E0717	Dyspnea	A071102	Failure to Interrogate	27/06/2022	27/06/2022
										A071204	No Pacing	27/06/2022	27/06/2022
								E2403	No Clinical Signs, Symptoms or Conditions	A071102	Failure to Interrogate	27/06/2022	27/06/2022
										A071204	No Pacing	27/06/2022	27/06/2022
								F1901	Additional Surgery	A071102	Failure to Interrogate	27/06/2022	27/06/2022
										A071204	No Pacing	27/06/2022	27/06/2022
						CT3002	Implantable synthetic polymer	E011903	Syncope/Fainting	A071102	Failure to Interrogate	27/06/2022	27/06/2022
										A071204	No Pacing	27/06/2022	27/06/2022

							E0601	Arrhythmia	A071102	Failure to Interrogate	27/06/2022	27/06/2022
									A071204	No Pacing	27/06/2022	27/06/2022
							E0717	Dyspnea	A071102	Failure to Interrogate	27/06/2022	27/06/2022
									A071204	No Pacing	27/06/2022	27/06/2022
							E2403	No Clinical Signs, Symptoms or Conditions	A071102	Failure to Interrogate	27/06/2022	27/06/2022
									A071204	No Pacing	27/06/2022	27/06/2022
							F1901	Additional Surgery	A071102	Failure to Interrogate	27/06/2022	27/06/2022
									A071204	No Pacing	27/06/2022	27/06/2022
					CT3006	Implantable inorganic material	E011903	Syncope/Fainting	A071102	Failure to Interrogate	27/06/2022	27/06/2022
									A071204	No Pacing	27/06/2022	27/06/2022
							E0601	Arrhythmia	A071102	Failure to Interrogate	27/06/2022	27/06/2022
									A071204	No Pacing	27/06/2022	27/06/2022
							E0717	Dyspnea	A071102	Failure to Interrogate	27/06/2022	27/06/2022
									A071204	No Pacing	27/06/2022	27/06/2022
							E2403	No Clinical Signs, Symptoms or Conditions	A071102	Failure to Interrogate	27/06/2022	27/06/2022
									A071204	No Pacing	27/06/2022	27/06/2022
							F1901	Additional Surgery	A071102	Failure to Interrogate	27/06/2022	27/06/2022
									A071204	No Pacing	27/06/2022	27/06/2022
					CT314	Electrically powered	E011903	Syncope/Fainting	A071102	Failure to Interrogate	27/06/2022	27/06/2022
									A071204	No Pacing	27/06/2022	27/06/2022
							E0601	Arrhythmia	A071102	Failure to Interrogate	27/06/2022	27/06/2022
									A071204	No Pacing	27/06/2022	27/06/2022
							E0717	Dyspnea	A071102	Failure to Interrogate	27/06/2022	27/06/2022
									A071204	No Pacing	27/06/2022	27/06/2022
							E2403	No Clinical Signs, Symptoms or Conditions	A071102	Failure to Interrogate	27/06/2022	27/06/2022
									A071204	No Pacing	27/06/2022	27/06/2022
							F1901	Additional Surgery	A071102	Failure to Interrogate	27/06/2022	27/06/2022
									A071204	No Pacing	27/06/2022	27/06/2022
					CT321	Long-term surgical invasive	E011903	Syncope/Fainting	A071102	Failure to Interrogate	27/06/2022	27/06/2022
									A071204	No Pacing	27/06/2022	27/06/2022
							E0601	Arrhythmia	A071102	Failure to Interrogate	27/06/2022	27/06/2022
									A071204	No Pacing	27/06/2022	27/06/2022
							E0717	Dyspnea	A071102	Failure to Interrogate	27/06/2022	27/06/2022
									A071204	No Pacing	27/06/2022	27/06/2022
							E2403	No Clinical Signs, Symptoms or Conditions	A071102	Failure to Interrogate	27/06/2022	27/06/2022
									A071204	No Pacing	27/06/2022	27/06/2022
							F1901	Additional Surgery	A071102	Failure to Interrogate	27/06/2022	27/06/2022
									A071204	No Pacing	27/06/2022	27/06/2022
					CT449	Generators and associated devices	E011903	Syncope/Fainting	A071102	Failure to Interrogate	27/06/2022	27/06/2022
									A071204	No Pacing	27/06/2022	27/06/2022
							E0601	Arrhythmia	A071102	Failure to Interrogate	27/06/2022	27/06/2022
									A071204	No Pacing	27/06/2022	27/06/2022
							E0717	Dyspnea	A071102	Failure to Interrogate	27/06/2022	27/06/2022
									A071204	No Pacing	27/06/2022	27/06/2022
							E2403	No Clinical Signs, Symptoms or Conditions	A071102	Failure to Interrogate	27/06/2022	27/06/2022
									A071204	No Pacing	27/06/2022	27/06/2022
							F1901	Additional Surgery	A071102	Failure to Interrogate	27/06/2022	27/06/2022
									A071204	No Pacing	27/06/2022	27/06/2022
					CT981	Single use	E011903	Syncope/Fainting	A071102	Failure to Interrogate	27/06/2022	27/06/2022
									A071204	No Pacing	27/06/2022	27/06/2022
							E0601	Arrhythmia	A071102	Failure to Interrogate	27/06/2022	27/06/2022
									A071204	No Pacing	27/06/2022	27/06/2022
							E0717	Dyspnea	A071102	Failure to Interrogate	27/06/2022	27/06/2022
									A071204	No Pacing	27/06/2022	27/06/2022
							E2403	No Clinical Signs, Symptoms or Conditions	A071102	Failure to Interrogate	27/06/2022	27/06/2022
									A071204	No Pacing	27/06/2022	27/06/2022
							F1901	Additional Surgery	A071102	Failure to Interrogate	27/06/2022	27/06/2022
									A071204	No Pacing	27/06/2022	27/06/2022
					CT983	Surgical invasive	E011903	Syncope/Fainting	A071102	Failure to Interrogate	27/06/2022	27/06/2022
									A071204	No Pacing	27/06/2022	27/06/2022
							E0601	Arrhythmia	A071102	Failure to Interrogate	27/06/2022	27/06/2022
									A071204	No Pacing	27/06/2022	27/06/2022
							E0717	Dyspnea	A071102	Failure to Interrogate	27/06/2022	27/06/2022
									A071204	No Pacing	27/06/2022	27/06/2022
							E2403	No Clinical Signs, Symptoms or Conditions	A071102	Failure to Interrogate	27/06/2022	27/06/2022
									A071204	No Pacing	27/06/2022	27/06/2022
							F1901	Additional Surgery	A071102	Failure to Interrogate	27/06/2022	27/06/2022
									A071204	No Pacing	27/06/2022	27/06/2022
7095	2022	All other reportable incidents	35223	Endocardial/interventricular septal pacing lead	CT1684	Cardiac pacemakers and associated devices	E0204	Lethargy	A0701	Capturing Problem		
									A070101	Failure to Capture		
									A070103	Intermittent Capture		
									A070910	Undersensing		
							E0601	Arrhythmia	A0701	Capturing Problem		
									A070101	Failure to Capture		
									A070103	Intermittent Capture		
									A070910	Undersensing		
							E060109	Tachycardia	A0701	Capturing Problem		
									A070101	Failure to Capture		
									A070103	Intermittent Capture		
									A070910	Undersensing		
							E0717	Dyspnea	A0701	Capturing Problem		
									A070101	Failure to Capture		
									A070103	Intermittent Capture		
									A070910	Undersensing		
							E2403	No Clinical Signs, Symptoms or Conditions	A0701	Capturing Problem		
									A070101	Failure to Capture		
									A070103	Intermittent Capture		
									A070910	Undersensing		

							F11	Non-Serious Injury/ Illness/ Impairment	A0701	Capturing Problem		
									A070101	Failure to Capture		
									A070103	Intermittent Capture		
									A070910	Undersensing		
7096	2021	All other reportable incidents	47270	Cardiac resynchronization therapy implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	11/11/2021	11/11/2021
									A071102	Failure to Interrogate	11/11/2021	11/11/2021
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery	11/11/2021	11/11/2021
									A071102	Failure to Interrogate	11/11/2021	11/11/2021
7097	2022	All other reportable incidents	47265	Dual-chamber implantable pacemaker, rate-responsive	CT141	Electrodes/Leads and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem	06/12/2022	06/12/2022
									A070504	Premature Discharge of Battery	06/12/2022	06/12/2022
							F08	Hospitalization or Prolonged Hospitalization	A0705	Battery Problem	06/12/2022	06/12/2022
									A070504	Premature Discharge of Battery	06/12/2022	06/12/2022
							F1905	Device Revision or Replacement	A0705	Battery Problem	06/12/2022	06/12/2022
									A070504	Premature Discharge of Battery	06/12/2022	06/12/2022
					CT145	Cardiovascular devices	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem	06/12/2022	06/12/2022
									A070504	Premature Discharge of Battery	06/12/2022	06/12/2022
							F08	Hospitalization or Prolonged Hospitalization	A0705	Battery Problem	06/12/2022	06/12/2022
									A070504	Premature Discharge of Battery	06/12/2022	06/12/2022
							F1905	Device Revision or Replacement	A0705	Battery Problem	06/12/2022	06/12/2022
									A070504	Premature Discharge of Battery	06/12/2022	06/12/2022
					CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem	06/12/2022	06/12/2022
									A070504	Premature Discharge of Battery	06/12/2022	06/12/2022
							F08	Hospitalization or Prolonged Hospitalization	A0705	Battery Problem	06/12/2022	06/12/2022
									A070504	Premature Discharge of Battery	06/12/2022	06/12/2022
							F1905	Device Revision or Replacement	A0705	Battery Problem	06/12/2022	06/12/2022
									A070504	Premature Discharge of Battery	06/12/2022	06/12/2022
					CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem	06/12/2022	06/12/2022
									A070504	Premature Discharge of Battery	06/12/2022	06/12/2022
							F08	Hospitalization or Prolonged Hospitalization	A0705	Battery Problem	06/12/2022	06/12/2022
									A070504	Premature Discharge of Battery	06/12/2022	06/12/2022
							F1905	Device Revision or Replacement	A0705	Battery Problem	06/12/2022	06/12/2022
									A070504	Premature Discharge of Battery	06/12/2022	06/12/2022
					CT233	Surgical	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem	06/12/2022	06/12/2022
									A070504	Premature Discharge of Battery	06/12/2022	06/12/2022
							F08	Hospitalization or Prolonged Hospitalization	A0705	Battery Problem	06/12/2022	06/12/2022
									A070504	Premature Discharge of Battery	06/12/2022	06/12/2022
							F1905	Device Revision or Replacement	A0705	Battery Problem	06/12/2022	06/12/2022
									A070504	Premature Discharge of Battery	06/12/2022	06/12/2022
					CT2406	Implantable	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem	06/12/2022	06/12/2022
									A070504	Premature Discharge of Battery	06/12/2022	06/12/2022
							F08	Hospitalization or Prolonged Hospitalization	A0705	Battery Problem	06/12/2022	06/12/2022
									A070504	Premature Discharge of Battery	06/12/2022	06/12/2022
							F1905	Device Revision or Replacement	A0705	Battery Problem	06/12/2022	06/12/2022
									A070504	Premature Discharge of Battery	06/12/2022	06/12/2022
					CT2407	Active implantable	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem	06/12/2022	06/12/2022
									A070504	Premature Discharge of Battery	06/12/2022	06/12/2022
							F08	Hospitalization or Prolonged Hospitalization	A0705	Battery Problem	06/12/2022	06/12/2022
									A070504	Premature Discharge of Battery	06/12/2022	06/12/2022
							F1905	Device Revision or Replacement	A0705	Battery Problem	06/12/2022	06/12/2022
									A070504	Premature Discharge of Battery	06/12/2022	06/12/2022
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem	06/12/2022	06/12/2022
									A070504	Premature Discharge of Battery	06/12/2022	06/12/2022
							F08	Hospitalization or Prolonged Hospitalization	A0705	Battery Problem	06/12/2022	06/12/2022
									A070504	Premature Discharge of Battery	06/12/2022	06/12/2022
							F1905	Device Revision or Replacement	A0705	Battery Problem	06/12/2022	06/12/2022
									A070504	Premature Discharge of Battery	06/12/2022	06/12/2022
					CT2727	Active implant devices/ancillaries	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem	06/12/2022	06/12/2022
									A070504	Premature Discharge of Battery	06/12/2022	06/12/2022
							F08	Hospitalization or Prolonged Hospitalization	A0705	Battery Problem	06/12/2022	06/12/2022
									A070504	Premature Discharge of Battery	06/12/2022	06/12/2022
							F1905	Device Revision or Replacement	A0705	Battery Problem	06/12/2022	06/12/2022
									A070504	Premature Discharge of Battery	06/12/2022	06/12/2022
					CT2750	Implantable cardiac electrical stimulators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem	06/12/2022	06/12/2022
									A070504	Premature Discharge of Battery	06/12/2022	06/12/2022
							F08	Hospitalization or Prolonged Hospitalization	A0705	Battery Problem	06/12/2022	06/12/2022
									A070504	Premature Discharge of Battery	06/12/2022	06/12/2022
							F1905	Device Revision or Replacement	A0705	Battery Problem	06/12/2022	06/12/2022
									A070504	Premature Discharge of Battery	06/12/2022	06/12/2022
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem	06/12/2022	06/12/2022
									A070504	Premature Discharge of Battery	06/12/2022	06/12/2022
							F08	Hospitalization or Prolonged Hospitalization	A0705	Battery Problem	06/12/2022	06/12/2022
									A070504	Premature Discharge of Battery	06/12/2022	06/12/2022
							F1905	Device Revision or Replacement	A0705	Battery Problem	06/12/2022	06/12/2022
									A070504	Premature Discharge of Battery	06/12/2022	06/12/2022
					CT2985	Active	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem	06/12/2022	06/12/2022
									A070504	Premature Discharge of Battery	06/12/2022	06/12/2022
							F08	Hospitalization or Prolonged Hospitalization	A0705	Battery Problem	06/12/2022	06/12/2022
									A070504	Premature Discharge of Battery	06/12/2022	06/12/2022
							F1905	Device Revision or Replacement	A0705	Battery Problem	06/12/2022	06/12/2022
									A070504	Premature Discharge of Battery	06/12/2022	06/12/2022
					CT3001	Implantable metal	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem	06/12/2022	06/12/2022
									A070504	Premature Discharge of Battery	06/12/2022	06/12/2022
							F08	Hospitalization or Prolonged Hospitalization	A0705	Battery Problem	06/12/2022	06/12/2022
									A070504	Premature Discharge of Battery	06/12/2022	06/12/2022
							F1905	Device Revision or Replacement	A0705	Battery Problem	06/12/2022	06/12/2022
									A070504	Premature Discharge of Battery	06/12/2022	06/12/2022
					CT3002	Implantable synthetic polymer	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem	06/12/2022	06/12/2022
									A070504	Premature Discharge of Battery	06/12/2022	06/12/2022

							F08	Hospitalization or Prolonged Hospitalization	A0705	Battery Problem	06/12/2022	06/12/2022
									A070504	Premature Discharge of Battery	06/12/2022	06/12/2022
							F1905	Device Revision or Replacement	A0705	Battery Problem	06/12/2022	06/12/2022
									A070504	Premature Discharge of Battery	06/12/2022	06/12/2022
					CT3006	Implantable inorganic material	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem	06/12/2022	06/12/2022
									A070504	Premature Discharge of Battery	06/12/2022	06/12/2022
							F08	Hospitalization or Prolonged Hospitalization	A0705	Battery Problem	06/12/2022	06/12/2022
									A070504	Premature Discharge of Battery	06/12/2022	06/12/2022
							F1905	Device Revision or Replacement	A0705	Battery Problem	06/12/2022	06/12/2022
									A070504	Premature Discharge of Battery	06/12/2022	06/12/2022
					CT314	Electrically powered	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem	06/12/2022	06/12/2022
									A070504	Premature Discharge of Battery	06/12/2022	06/12/2022
							F08	Hospitalization or Prolonged Hospitalization	A0705	Battery Problem	06/12/2022	06/12/2022
									A070504	Premature Discharge of Battery	06/12/2022	06/12/2022
							F1905	Device Revision or Replacement	A0705	Battery Problem	06/12/2022	06/12/2022
									A070504	Premature Discharge of Battery	06/12/2022	06/12/2022
					CT321	Long-term surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem	06/12/2022	06/12/2022
									A070504	Premature Discharge of Battery	06/12/2022	06/12/2022
							F08	Hospitalization or Prolonged Hospitalization	A0705	Battery Problem	06/12/2022	06/12/2022
									A070504	Premature Discharge of Battery	06/12/2022	06/12/2022
							F1905	Device Revision or Replacement	A0705	Battery Problem	06/12/2022	06/12/2022
									A070504	Premature Discharge of Battery	06/12/2022	06/12/2022
					CT449	Generators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem	06/12/2022	06/12/2022
									A070504	Premature Discharge of Battery	06/12/2022	06/12/2022
							F08	Hospitalization or Prolonged Hospitalization	A0705	Battery Problem	06/12/2022	06/12/2022
									A070504	Premature Discharge of Battery	06/12/2022	06/12/2022
							F1905	Device Revision or Replacement	A0705	Battery Problem	06/12/2022	06/12/2022
									A070504	Premature Discharge of Battery	06/12/2022	06/12/2022
					CT981	Single use	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem	06/12/2022	06/12/2022
									A070504	Premature Discharge of Battery	06/12/2022	06/12/2022
							F08	Hospitalization or Prolonged Hospitalization	A0705	Battery Problem	06/12/2022	06/12/2022
									A070504	Premature Discharge of Battery	06/12/2022	06/12/2022
							F1905	Device Revision or Replacement	A0705	Battery Problem	06/12/2022	06/12/2022
									A070504	Premature Discharge of Battery	06/12/2022	06/12/2022
					CT983	Surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem	06/12/2022	06/12/2022
									A070504	Premature Discharge of Battery	06/12/2022	06/12/2022
							F08	Hospitalization or Prolonged Hospitalization	A0705	Battery Problem	06/12/2022	06/12/2022
									A070504	Premature Discharge of Battery	06/12/2022	06/12/2022
							F1905	Device Revision or Replacement	A0705	Battery Problem	06/12/2022	06/12/2022
									A070504	Premature Discharge of Battery	06/12/2022	06/12/2022
7098	2022	All other reportable incidents	17882	Physiologic-monitoring defibrillation system	CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A07	Electrical /Electronic Property Problem		
							F27	Problem Identified During Non-Clinical Procedure	A07	Electrical /Electronic Property Problem		
7099	2021	All other reportable incidents	47265	Dual-chamber implantable pacemaker, rate-responsive	CT141	Electrodes/Leads and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	05/01/2022	05/01/2022
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	05/01/2022	05/01/2022
							F1903	Device Explantation	A070504	Premature Discharge of Battery	05/01/2022	05/01/2022
					CT145	Cardiovascular devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	05/01/2022	05/01/2022
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	05/01/2022	05/01/2022
							F1903	Device Explantation	A070504	Premature Discharge of Battery	05/01/2022	05/01/2022
					CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	05/01/2022	05/01/2022
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	05/01/2022	05/01/2022
							F1903	Device Explantation	A070504	Premature Discharge of Battery	05/01/2022	05/01/2022
					CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	05/01/2022	05/01/2022
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	05/01/2022	05/01/2022
							F1903	Device Explantation	A070504	Premature Discharge of Battery	05/01/2022	05/01/2022
					CT233	Surgical	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	05/01/2022	05/01/2022
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	05/01/2022	05/01/2022
							F1903	Device Explantation	A070504	Premature Discharge of Battery	05/01/2022	05/01/2022
					CT2406	Implantable	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	05/01/2022	05/01/2022
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	05/01/2022	05/01/2022
							F1903	Device Explantation	A070504	Premature Discharge of Battery	05/01/2022	05/01/2022
					CT2407	Active implantable	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	05/01/2022	05/01/2022
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	05/01/2022	05/01/2022
							F1903	Device Explantation	A070504	Premature Discharge of Battery	05/01/2022	05/01/2022
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	05/01/2022	05/01/2022
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	05/01/2022	05/01/2022
							F1903	Device Explantation	A070504	Premature Discharge of Battery	05/01/2022	05/01/2022
					CT2727	Active implant devices/ancillaries	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	05/01/2022	05/01/2022
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	05/01/2022	05/01/2022
							F1903	Device Explantation	A070504	Premature Discharge of Battery	05/01/2022	05/01/2022
					CT2750	Implantable cardiac electrical stimulators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	05/01/2022	05/01/2022
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	05/01/2022	05/01/2022
							F1903	Device Explantation	A070504	Premature Discharge of Battery	05/01/2022	05/01/2022
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	05/01/2022	05/01/2022
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	05/01/2022	05/01/2022
							F1903	Device Explantation	A070504	Premature Discharge of Battery	05/01/2022	05/01/2022
					CT2985	Active	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	05/01/2022	05/01/2022
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	05/01/2022	05/01/2022
							F1903	Device Explantation	A070504	Premature Discharge of Battery	05/01/2022	05/01/2022
					CT3001	Implantable metal	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	05/01/2022	05/01/2022
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	05/01/2022	05/01/2022
							F1903	Device Explantation	A070504	Premature Discharge of Battery	05/01/2022	05/01/2022
					CT3002	Implantable synthetic polymer	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	05/01/2022	05/01/2022
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	05/01/2022	05/01/2022
							F1903	Device Explantation	A070504	Premature Discharge of Battery	05/01/2022	05/01/2022
					CT3006	Implantable inorganic material	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	05/01/2022	05/01/2022
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	05/01/2022	05/01/2022
							F1903	Device Explantation	A070504	Premature Discharge of Battery	05/01/2022	05/01/2022
					CT314	Electrically powered	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	05/01/2022	05/01/2022

							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	05/01/2022	05/01/2022
							F1903	Device Explantation	A070504	Premature Discharge of Battery	05/01/2022	05/01/2022
					CT321	Long-term surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	05/01/2022	05/01/2022
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	05/01/2022	05/01/2022
							F1903	Device Explantation	A070504	Premature Discharge of Battery	05/01/2022	05/01/2022
					CT449	Generators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	05/01/2022	05/01/2022
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	05/01/2022	05/01/2022
							F1903	Device Explantation	A070504	Premature Discharge of Battery	05/01/2022	05/01/2022
					CT981	Single use	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	05/01/2022	05/01/2022
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	05/01/2022	05/01/2022
							F1903	Device Explantation	A070504	Premature Discharge of Battery	05/01/2022	05/01/2022
					CT983	Surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	05/01/2022	05/01/2022
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	05/01/2022	05/01/2022
							F1903	Device Explantation	A070504	Premature Discharge of Battery	05/01/2022	05/01/2022
7100	2021	All other reportable incidents	47265	Dual-chamber implantable pacemaker, rate-responsive	CT141	Electrodes/Leads and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A0709	Device Sensing Problem	14/07/2021	14/07/2021
									A0712	Pacing Problem	14/07/2021	14/07/2021
									A072201	High Impedance	14/07/2021	14/07/2021
							F1908	Prolonged Surgery	A0709	Device Sensing Problem	14/07/2021	14/07/2021
									A0712	Pacing Problem	14/07/2021	14/07/2021
									A072201	High Impedance	14/07/2021	14/07/2021
					CT145	Cardiovascular devices	E2403	No Clinical Signs, Symptoms or Conditions	A0709	Device Sensing Problem	14/07/2021	14/07/2021
									A0712	Pacing Problem	14/07/2021	14/07/2021
									A072201	High Impedance	14/07/2021	14/07/2021
							F1908	Prolonged Surgery	A0709	Device Sensing Problem	14/07/2021	14/07/2021
									A0712	Pacing Problem	14/07/2021	14/07/2021
									A072201	High Impedance	14/07/2021	14/07/2021
					CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A0709	Device Sensing Problem	14/07/2021	14/07/2021
									A0712	Pacing Problem	14/07/2021	14/07/2021
									A072201	High Impedance	14/07/2021	14/07/2021
							F1908	Prolonged Surgery	A0709	Device Sensing Problem	14/07/2021	14/07/2021
									A0712	Pacing Problem	14/07/2021	14/07/2021
									A072201	High Impedance	14/07/2021	14/07/2021
					CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A0709	Device Sensing Problem	14/07/2021	14/07/2021
									A0712	Pacing Problem	14/07/2021	14/07/2021
									A072201	High Impedance	14/07/2021	14/07/2021
							F1908	Prolonged Surgery	A0709	Device Sensing Problem	14/07/2021	14/07/2021
									A0712	Pacing Problem	14/07/2021	14/07/2021
									A072201	High Impedance	14/07/2021	14/07/2021
					CT233	Surgical	E2403	No Clinical Signs, Symptoms or Conditions	A0709	Device Sensing Problem	14/07/2021	14/07/2021
									A0712	Pacing Problem	14/07/2021	14/07/2021
									A072201	High Impedance	14/07/2021	14/07/2021
							F1908	Prolonged Surgery	A0709	Device Sensing Problem	14/07/2021	14/07/2021
									A0712	Pacing Problem	14/07/2021	14/07/2021
									A072201	High Impedance	14/07/2021	14/07/2021
					CT2406	Implantable	E2403	No Clinical Signs, Symptoms or Conditions	A0709	Device Sensing Problem	14/07/2021	14/07/2021
									A0712	Pacing Problem	14/07/2021	14/07/2021
									A072201	High Impedance	14/07/2021	14/07/2021
							F1908	Prolonged Surgery	A0709	Device Sensing Problem	14/07/2021	14/07/2021
									A0712	Pacing Problem	14/07/2021	14/07/2021
									A072201	High Impedance	14/07/2021	14/07/2021
					CT2407	Active implantable	E2403	No Clinical Signs, Symptoms or Conditions	A0709	Device Sensing Problem	14/07/2021	14/07/2021
									A0712	Pacing Problem	14/07/2021	14/07/2021
									A072201	High Impedance	14/07/2021	14/07/2021
							F1908	Prolonged Surgery	A0709	Device Sensing Problem	14/07/2021	14/07/2021
									A0712	Pacing Problem	14/07/2021	14/07/2021
									A072201	High Impedance	14/07/2021	14/07/2021
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A0709	Device Sensing Problem	14/07/2021	14/07/2021
									A0712	Pacing Problem	14/07/2021	14/07/2021
									A072201	High Impedance	14/07/2021	14/07/2021
							F1908	Prolonged Surgery	A0709	Device Sensing Problem	14/07/2021	14/07/2021
									A0712	Pacing Problem	14/07/2021	14/07/2021
									A072201	High Impedance	14/07/2021	14/07/2021
					CT2727	Active implant devices/ancillaries	E2403	No Clinical Signs, Symptoms or Conditions	A0709	Device Sensing Problem	14/07/2021	14/07/2021
									A0712	Pacing Problem	14/07/2021	14/07/2021
									A072201	High Impedance	14/07/2021	14/07/2021
							F1908	Prolonged Surgery	A0709	Device Sensing Problem	14/07/2021	14/07/2021
									A0712	Pacing Problem	14/07/2021	14/07/2021
									A072201	High Impedance	14/07/2021	14/07/2021
					CT2750	Implantable cardiac electrical stimulators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A0709	Device Sensing Problem	14/07/2021	14/07/2021
									A0712	Pacing Problem	14/07/2021	14/07/2021
									A072201	High Impedance	14/07/2021	14/07/2021
							F1908	Prolonged Surgery	A0709	Device Sensing Problem	14/07/2021	14/07/2021
									A0712	Pacing Problem	14/07/2021	14/07/2021
									A072201	High Impedance	14/07/2021	14/07/2021
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A0709	Device Sensing Problem	14/07/2021	14/07/2021
									A0712	Pacing Problem	14/07/2021	14/07/2021
									A072201	High Impedance	14/07/2021	14/07/2021
							F1908	Prolonged Surgery	A0709	Device Sensing Problem	14/07/2021	14/07/2021
									A0712	Pacing Problem	14/07/2021	14/07/2021
									A072201	High Impedance	14/07/2021	14/07/2021
					CT2985	Active	E2403	No Clinical Signs, Symptoms or Conditions	A0709	Device Sensing Problem	14/07/2021	14/07/2021
									A0712	Pacing Problem	14/07/2021	14/07/2021
									A072201	High Impedance	14/07/2021	14/07/2021
							F1908	Prolonged Surgery	A0709	Device Sensing Problem	14/07/2021	14/07/2021
									A0712	Pacing Problem	14/07/2021	14/07/2021
									A072201	High Impedance	14/07/2021	14/07/2021
					CT3001	Implantable metal	E2403	No Clinical Signs, Symptoms or Conditions	A0709	Device Sensing Problem	14/07/2021	14/07/2021
									A0712	Pacing Problem	14/07/2021	14/07/2021

								A072201	High Impedance	14/07/2021	14/07/2021	
						F1908	Prolonged Surgery	A0709	Device Sensing Problem	14/07/2021	14/07/2021	
								A0712	Pacing Problem	14/07/2021	14/07/2021	
								A072201	High Impedance	14/07/2021	14/07/2021	
					CT3002	Implantable synthetic polymer	E2403	No Clinical Signs, Symptoms or Conditions	A0709	Device Sensing Problem	14/07/2021	14/07/2021
									A0712	Pacing Problem	14/07/2021	14/07/2021
									A072201	High Impedance	14/07/2021	14/07/2021
							F1908	Prolonged Surgery	A0709	Device Sensing Problem	14/07/2021	14/07/2021
									A0712	Pacing Problem	14/07/2021	14/07/2021
									A072201	High Impedance	14/07/2021	14/07/2021
					CT3006	Implantable inorganic material	E2403	No Clinical Signs, Symptoms or Conditions	A0709	Device Sensing Problem	14/07/2021	14/07/2021
									A0712	Pacing Problem	14/07/2021	14/07/2021
									A072201	High Impedance	14/07/2021	14/07/2021
							F1908	Prolonged Surgery	A0709	Device Sensing Problem	14/07/2021	14/07/2021
									A0712	Pacing Problem	14/07/2021	14/07/2021
									A072201	High Impedance	14/07/2021	14/07/2021
					CT314	Electrically powered	E2403	No Clinical Signs, Symptoms or Conditions	A0709	Device Sensing Problem	14/07/2021	14/07/2021
									A0712	Pacing Problem	14/07/2021	14/07/2021
									A072201	High Impedance	14/07/2021	14/07/2021
							F1908	Prolonged Surgery	A0709	Device Sensing Problem	14/07/2021	14/07/2021
									A0712	Pacing Problem	14/07/2021	14/07/2021
									A072201	High Impedance	14/07/2021	14/07/2021
					CT321	Long-term surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A0709	Device Sensing Problem	14/07/2021	14/07/2021
									A0712	Pacing Problem	14/07/2021	14/07/2021
									A072201	High Impedance	14/07/2021	14/07/2021
							F1908	Prolonged Surgery	A0709	Device Sensing Problem	14/07/2021	14/07/2021
									A0712	Pacing Problem	14/07/2021	14/07/2021
									A072201	High Impedance	14/07/2021	14/07/2021
					CT449	Generators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A0709	Device Sensing Problem	14/07/2021	14/07/2021
									A0712	Pacing Problem	14/07/2021	14/07/2021
									A072201	High Impedance	14/07/2021	14/07/2021
							F1908	Prolonged Surgery	A0709	Device Sensing Problem	14/07/2021	14/07/2021
									A0712	Pacing Problem	14/07/2021	14/07/2021
									A072201	High Impedance	14/07/2021	14/07/2021
					CT981	Single use	E2403	No Clinical Signs, Symptoms or Conditions	A0709	Device Sensing Problem	14/07/2021	14/07/2021
									A0712	Pacing Problem	14/07/2021	14/07/2021
									A072201	High Impedance	14/07/2021	14/07/2021
							F1908	Prolonged Surgery	A0709	Device Sensing Problem	14/07/2021	14/07/2021
									A0712	Pacing Problem	14/07/2021	14/07/2021
									A072201	High Impedance	14/07/2021	14/07/2021
					CT983	Surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A0709	Device Sensing Problem	14/07/2021	14/07/2021
									A0712	Pacing Problem	14/07/2021	14/07/2021
									A072201	High Impedance	14/07/2021	14/07/2021
							F1908	Prolonged Surgery	A0709	Device Sensing Problem	14/07/2021	14/07/2021
									A0712	Pacing Problem	14/07/2021	14/07/2021
									A072201	High Impedance	14/07/2021	14/07/2021
7101	2021	All other reportable incidents	35852	Single-chamber implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A072201	High Impedance		
							F26	No Health Consequences or Impact	A072201	High Impedance		
7102	2021	All other reportable incidents	60789	Intracardiac pacemaker	CT1685	Cardiac pacemakers	E0605	Cardiac Tamponade	A24	No Apparent Device Problem		
							E0619	Pericardial Effusion	A24	No Apparent Device Problem		
							E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem		
							F08	Hospitalization or Prolonged Hospitalization	A24	No Apparent Device Problem		
							F1203	Life Threatening Illness or Injury	A24	No Apparent Device Problem		
							F23	Unexpected Medical Intervention	A24	No Apparent Device Problem		
					CT257	Implantable cardiac pacemakers	E0605	Cardiac Tamponade	A24	No Apparent Device Problem		
							E0619	Pericardial Effusion	A24	No Apparent Device Problem		
							E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem		
							F08	Hospitalization or Prolonged Hospitalization	A24	No Apparent Device Problem		
							F1203	Life Threatening Illness or Injury	A24	No Apparent Device Problem		
							F23	Unexpected Medical Intervention	A24	No Apparent Device Problem		
					CT2818	Catheterization/Cannulation/Fluid path devices	E0605	Cardiac Tamponade	A24	No Apparent Device Problem		
							E0619	Pericardial Effusion	A24	No Apparent Device Problem		
							E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem		
							F08	Hospitalization or Prolonged Hospitalization	A24	No Apparent Device Problem		
							F1203	Life Threatening Illness or Injury	A24	No Apparent Device Problem		
							F23	Unexpected Medical Intervention	A24	No Apparent Device Problem		
					CT2866	Implantable cardiac electrical stimulators	E0605	Cardiac Tamponade	A24	No Apparent Device Problem		
							E0619	Pericardial Effusion	A24	No Apparent Device Problem		
							E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem		
							F08	Hospitalization or Prolonged Hospitalization	A24	No Apparent Device Problem		
							F1203	Life Threatening Illness or Injury	A24	No Apparent Device Problem		
							F23	Unexpected Medical Intervention	A24	No Apparent Device Problem		
7103	2021	Death	47265	Dual-chamber implantable pacemaker, rate-responsive	CT141	Electrodes/Leads and associated devices	E2402	Appropriate Clinical Signs, Symptoms and Conditions Term/Code not Available	A1107	Data Problem		
							F02	Death	A1107	Data Problem		
					CT145	Cardiovascular devices	E2402	Appropriate Clinical Signs, Symptoms and Conditions Term/Code not Available	A1107	Data Problem		
							F02	Death	A1107	Data Problem		
					CT1684	Cardiac pacemakers and associated devices	E2402	Appropriate Clinical Signs, Symptoms and Conditions Term/Code not Available	A1107	Data Problem		
							F02	Death	A1107	Data Problem		
					CT1685	Cardiac pacemakers	E2402	Appropriate Clinical Signs, Symptoms and Conditions Term/Code not Available	A1107	Data Problem		
							F02	Death	A1107	Data Problem		
					CT233	Surgical	E2402	Appropriate Clinical Signs, Symptoms and Conditions Term/Code not Available	A1107	Data Problem		
							F02	Death	A1107	Data Problem		
					CT2406	Implantable	E2402	Appropriate Clinical Signs, Symptoms and Conditions Term/Code not Available	A1107	Data Problem		

							F02	Death	A1107	Data Problem			
					CT2407	Active implantable	E2402	Appropriate Clinical Signs, Symptoms and Conditions Term/Code not Available	A1107	Data Problem			
							F02	Death	A1107	Data Problem			
					CT257	Implantable cardiac pacemakers	E2402	Appropriate Clinical Signs, Symptoms and Conditions Term/Code not Available	A1107	Data Problem			
							F02	Death	A1107	Data Problem			
					CT2727	Active implant devices/ancillaries	E2402	Appropriate Clinical Signs, Symptoms and Conditions Term/Code not Available	A1107	Data Problem			
							F02	Death	A1107	Data Problem			
					CT2750	Implantable cardiac electrical stimulators and associated devices	E2402	Appropriate Clinical Signs, Symptoms and Conditions Term/Code not Available	A1107	Data Problem			
							F02	Death	A1107	Data Problem			
					CT2866	Implantable cardiac electrical stimulators	E2402	Appropriate Clinical Signs, Symptoms and Conditions Term/Code not Available	A1107	Data Problem			
							F02	Death	A1107	Data Problem			
					CT2985	Active	E2402	Appropriate Clinical Signs, Symptoms and Conditions Term/Code not Available	A1107	Data Problem			
							F02	Death	A1107	Data Problem			
					CT3001	Implantable metal	E2402	Appropriate Clinical Signs, Symptoms and Conditions Term/Code not Available	A1107	Data Problem			
							F02	Death	A1107	Data Problem			
					CT3002	Implantable synthetic polymer	E2402	Appropriate Clinical Signs, Symptoms and Conditions Term/Code not Available	A1107	Data Problem			
							F02	Death	A1107	Data Problem			
					CT3006	Implantable inorganic material	E2402	Appropriate Clinical Signs, Symptoms and Conditions Term/Code not Available	A1107	Data Problem			
							F02	Death	A1107	Data Problem			
					CT314	Electrically powered	E2402	Appropriate Clinical Signs, Symptoms and Conditions Term/Code not Available	A1107	Data Problem			
							F02	Death	A1107	Data Problem			
					CT321	Long-term surgical invasive	E2402	Appropriate Clinical Signs, Symptoms and Conditions Term/Code not Available	A1107	Data Problem			
							F02	Death	A1107	Data Problem			
					CT449	Generators and associated devices	E2402	Appropriate Clinical Signs, Symptoms and Conditions Term/Code not Available	A1107	Data Problem			
							F02	Death	A1107	Data Problem			
					CT981	Single use	E2402	Appropriate Clinical Signs, Symptoms and Conditions Term/Code not Available	A1107	Data Problem			
							F02	Death	A1107	Data Problem			
					CT983	Surgical invasive	E2402	Appropriate Clinical Signs, Symptoms and Conditions Term/Code not Available	A1107	Data Problem			
							F02	Death	A1107	Data Problem			
7104	2021	All other reportable incidents	47265	Dual-chamber implantable pacemaker, rate-responsive	CT141	Electrodes/Leads and associated devices	E011902	Presyncope	A070909	Oversensing			
									A0712	Pacing Problem			
									A072201	High Impedance			
							E2403	No Clinical Signs, Symptoms or Conditions	A070909	Oversensing			
									A0712	Pacing Problem			
									A072201	High Impedance			
							F1001	Absence of Treatment	A070909	Oversensing			
									A0712	Pacing Problem			
									A072201	High Impedance			
							F15	Recognized Device or Procedural Complication	A070909	Oversensing			
									A0712	Pacing Problem			
									A072201	High Impedance			
							F2203	Imaging Required	A070909	Oversensing			
									A0712	Pacing Problem			
									A072201	High Impedance			
					CT145	Cardiovascular devices	E011902	Presyncope	A070909	Oversensing			
									A0712	Pacing Problem			
									A072201	High Impedance			
							E2403	No Clinical Signs, Symptoms or Conditions	A070909	Oversensing			
									A0712	Pacing Problem			
									A072201	High Impedance			
							F1001	Absence of Treatment	A070909	Oversensing			
									A0712	Pacing Problem			
									A072201	High Impedance			
							F15	Recognized Device or Procedural Complication	A070909	Oversensing			
									A0712	Pacing Problem			
									A072201	High Impedance			
							F2203	Imaging Required	A070909	Oversensing			
									A0712	Pacing Problem			
									A072201	High Impedance			
					CT1684	Cardiac pacemakers and associated devices	E011902	Presyncope	A070909	Oversensing			
									A0712	Pacing Problem			
									A072201	High Impedance			
							E2403	No Clinical Signs, Symptoms or Conditions	A070909	Oversensing			
									A0712	Pacing Problem			
									A072201	High Impedance			
							F1001	Absence of Treatment	A070909	Oversensing			
									A0712	Pacing Problem			
									A072201	High Impedance			
							F15	Recognized Device or Procedural Complication	A070909	Oversensing			
									A0712	Pacing Problem			
									A072201	High Impedance			
							F2203	Imaging Required	A070909	Oversensing			
									A0712	Pacing Problem			
									A072201	High Impedance			
					CT1685	Cardiac pacemakers	E011902	Presyncope	A070909	Oversensing			
									A0712	Pacing Problem			
									A072201	High Impedance			

							E2403	No Clinical Signs, Symptoms or Conditions	A070909	Oversensing		
									A0712	Pacing Problem		
									A072201	High Impedance		
							F1001	Absence of Treatment	A070909	Oversensing		
									A0712	Pacing Problem		
									A072201	High Impedance		
							F15	Recognized Device or Procedural Complication	A070909	Oversensing		
									A0712	Pacing Problem		
									A072201	High Impedance		
							F2203	Imaging Required	A070909	Oversensing		
									A0712	Pacing Problem		
									A072201	High Impedance		
					CT233	Surgical	E011902	Presyncope	A070909	Oversensing		
									A0712	Pacing Problem		
									A072201	High Impedance		
							E2403	No Clinical Signs, Symptoms or Conditions	A070909	Oversensing		
									A0712	Pacing Problem		
									A072201	High Impedance		
							F1001	Absence of Treatment	A070909	Oversensing		
									A0712	Pacing Problem		
									A072201	High Impedance		
							F15	Recognized Device or Procedural Complication	A070909	Oversensing		
									A0712	Pacing Problem		
									A072201	High Impedance		
							F2203	Imaging Required	A070909	Oversensing		
									A0712	Pacing Problem		
									A072201	High Impedance		
					CT2406	Implantable	E011902	Presyncope	A070909	Oversensing		
									A0712	Pacing Problem		
									A072201	High Impedance		
							E2403	No Clinical Signs, Symptoms or Conditions	A070909	Oversensing		
									A0712	Pacing Problem		
									A072201	High Impedance		
							F1001	Absence of Treatment	A070909	Oversensing		
									A0712	Pacing Problem		
									A072201	High Impedance		
							F15	Recognized Device or Procedural Complication	A070909	Oversensing		
									A0712	Pacing Problem		
									A072201	High Impedance		
							F2203	Imaging Required	A070909	Oversensing		
									A0712	Pacing Problem		
									A072201	High Impedance		
					CT2407	Active implantable	E011902	Presyncope	A070909	Oversensing		
									A0712	Pacing Problem		
									A072201	High Impedance		
							E2403	No Clinical Signs, Symptoms or Conditions	A070909	Oversensing		
									A0712	Pacing Problem		
									A072201	High Impedance		
							F1001	Absence of Treatment	A070909	Oversensing		
									A0712	Pacing Problem		
									A072201	High Impedance		
							F15	Recognized Device or Procedural Complication	A070909	Oversensing		
									A0712	Pacing Problem		
									A072201	High Impedance		
							F2203	Imaging Required	A070909	Oversensing		
									A0712	Pacing Problem		
									A072201	High Impedance		
					CT257	Implantable cardiac pacemakers	E011902	Presyncope	A070909	Oversensing		
									A0712	Pacing Problem		
									A072201	High Impedance		
							E2403	No Clinical Signs, Symptoms or Conditions	A070909	Oversensing		
									A0712	Pacing Problem		
									A072201	High Impedance		
							F1001	Absence of Treatment	A070909	Oversensing		
									A0712	Pacing Problem		
									A072201	High Impedance		
							F15	Recognized Device or Procedural Complication	A070909	Oversensing		
									A0712	Pacing Problem		
									A072201	High Impedance		
							F2203	Imaging Required	A070909	Oversensing		
									A0712	Pacing Problem		
									A072201	High Impedance		
					CT2727	Active implant devices/ancillaries	E011902	Presyncope	A070909	Oversensing		
									A0712	Pacing Problem		
									A072201	High Impedance		
							E2403	No Clinical Signs, Symptoms or Conditions	A070909	Oversensing		
									A0712	Pacing Problem		
									A072201	High Impedance		
							F1001	Absence of Treatment	A070909	Oversensing		
									A0712	Pacing Problem		
									A072201	High Impedance		
							F15	Recognized Device or Procedural Complication	A070909	Oversensing		
									A0712	Pacing Problem		
									A072201	High Impedance		
							F2203	Imaging Required	A070909	Oversensing		
									A0712	Pacing Problem		
									A072201	High Impedance		
					CT2750	Implantable cardiac electrical stimulators and associated devices	E011902	Presyncope	A070909	Oversensing		

									A0712	Pacing Problem		
									A072201	High Impedance		
							E2403	No Clinical Signs, Symptoms or Conditions	A070909	Oversensing		
									A0712	Pacing Problem		
									A072201	High Impedance		
							F1001	Absence of Treatment	A070909	Oversensing		
									A0712	Pacing Problem		
									A072201	High Impedance		
							F15	Recognized Device or Procedural Complication	A070909	Oversensing		
									A0712	Pacing Problem		
									A072201	High Impedance		
							F2203	Imaging Required	A070909	Oversensing		
									A0712	Pacing Problem		
									A072201	High Impedance		
					CT2866	Implantable cardiac electrical stimulators	E011902	Presyncope	A070909	Oversensing		
									A0712	Pacing Problem		
									A072201	High Impedance		
							E2403	No Clinical Signs, Symptoms or Conditions	A070909	Oversensing		
									A0712	Pacing Problem		
									A072201	High Impedance		
							F1001	Absence of Treatment	A070909	Oversensing		
									A0712	Pacing Problem		
									A072201	High Impedance		
							F15	Recognized Device or Procedural Complication	A070909	Oversensing		
									A0712	Pacing Problem		
									A072201	High Impedance		
							F2203	Imaging Required	A070909	Oversensing		
									A0712	Pacing Problem		
									A072201	High Impedance		
					CT2985	Active	E011902	Presyncope	A070909	Oversensing		
									A0712	Pacing Problem		
									A072201	High Impedance		
							E2403	No Clinical Signs, Symptoms or Conditions	A070909	Oversensing		
									A0712	Pacing Problem		
									A072201	High Impedance		
							F1001	Absence of Treatment	A070909	Oversensing		
									A0712	Pacing Problem		
									A072201	High Impedance		
							F15	Recognized Device or Procedural Complication	A070909	Oversensing		
									A0712	Pacing Problem		
									A072201	High Impedance		
							F2203	Imaging Required	A070909	Oversensing		
									A0712	Pacing Problem		
									A072201	High Impedance		
					CT3001	Implantable metal	E011902	Presyncope	A070909	Oversensing		
									A0712	Pacing Problem		
									A072201	High Impedance		
							E2403	No Clinical Signs, Symptoms or Conditions	A070909	Oversensing		
									A0712	Pacing Problem		
									A072201	High Impedance		
							F1001	Absence of Treatment	A070909	Oversensing		
									A0712	Pacing Problem		
									A072201	High Impedance		
							F15	Recognized Device or Procedural Complication	A070909	Oversensing		
									A0712	Pacing Problem		
									A072201	High Impedance		
							F2203	Imaging Required	A070909	Oversensing		
									A0712	Pacing Problem		
									A072201	High Impedance		
					CT3002	Implantable synthetic polymer	E011902	Presyncope	A070909	Oversensing		
									A0712	Pacing Problem		
									A072201	High Impedance		
							E2403	No Clinical Signs, Symptoms or Conditions	A070909	Oversensing		
									A0712	Pacing Problem		
									A072201	High Impedance		
							F1001	Absence of Treatment	A070909	Oversensing		
									A0712	Pacing Problem		
									A072201	High Impedance		
							F15	Recognized Device or Procedural Complication	A070909	Oversensing		
									A0712	Pacing Problem		
									A072201	High Impedance		
							F2203	Imaging Required	A070909	Oversensing		
									A0712	Pacing Problem		
									A072201	High Impedance		
					CT3006	Implantable inorganic material	E011902	Presyncope	A070909	Oversensing		
									A0712	Pacing Problem		
									A072201	High Impedance		
							E2403	No Clinical Signs, Symptoms or Conditions	A070909	Oversensing		
									A0712	Pacing Problem		
									A072201	High Impedance		
							F1001	Absence of Treatment	A070909	Oversensing		
									A0712	Pacing Problem		
									A072201	High Impedance		
							F15	Recognized Device or Procedural Complication	A070909	Oversensing		
									A0712	Pacing Problem		
									A072201	High Impedance		
							F2203	Imaging Required	A070909	Oversensing		
									A0712	Pacing Problem		

									A072201	High Impedance		
					CT314	Electrically powered	E011902	Presyncope	A070909	Oversensing		
									A0712	Pacing Problem		
									A072201	High Impedance		
							E2403	No Clinical Signs, Symptoms or Conditions	A070909	Oversensing		
									A0712	Pacing Problem		
									A072201	High Impedance		
							F1001	Absence of Treatment	A070909	Oversensing		
									A0712	Pacing Problem		
									A072201	High Impedance		
							F15	Recognized Device or Procedural Complication	A070909	Oversensing		
									A0712	Pacing Problem		
									A072201	High Impedance		
							F2203	Imaging Required	A070909	Oversensing		
									A0712	Pacing Problem		
									A072201	High Impedance		
					CT321	Long-term surgical invasive	E011902	Presyncope	A070909	Oversensing		
									A0712	Pacing Problem		
									A072201	High Impedance		
							E2403	No Clinical Signs, Symptoms or Conditions	A070909	Oversensing		
									A0712	Pacing Problem		
									A072201	High Impedance		
							F1001	Absence of Treatment	A070909	Oversensing		
									A0712	Pacing Problem		
									A072201	High Impedance		
							F15	Recognized Device or Procedural Complication	A070909	Oversensing		
									A0712	Pacing Problem		
									A072201	High Impedance		
							F2203	Imaging Required	A070909	Oversensing		
									A0712	Pacing Problem		
									A072201	High Impedance		
					CT449	Generators and associated devices	E011902	Presyncope	A070909	Oversensing		
									A0712	Pacing Problem		
									A072201	High Impedance		
							E2403	No Clinical Signs, Symptoms or Conditions	A070909	Oversensing		
									A0712	Pacing Problem		
									A072201	High Impedance		
							F1001	Absence of Treatment	A070909	Oversensing		
									A0712	Pacing Problem		
									A072201	High Impedance		
							F15	Recognized Device or Procedural Complication	A070909	Oversensing		
									A0712	Pacing Problem		
									A072201	High Impedance		
							F2203	Imaging Required	A070909	Oversensing		
									A0712	Pacing Problem		
									A072201	High Impedance		
					CT981	Single use	E011902	Presyncope	A070909	Oversensing		
									A0712	Pacing Problem		
									A072201	High Impedance		
							E2403	No Clinical Signs, Symptoms or Conditions	A070909	Oversensing		
									A0712	Pacing Problem		
									A072201	High Impedance		
							F1001	Absence of Treatment	A070909	Oversensing		
									A0712	Pacing Problem		
									A072201	High Impedance		
							F15	Recognized Device or Procedural Complication	A070909	Oversensing		
									A0712	Pacing Problem		
									A072201	High Impedance		
							F2203	Imaging Required	A070909	Oversensing		
									A0712	Pacing Problem		
									A072201	High Impedance		
					CT983	Surgical invasive	E011902	Presyncope	A070909	Oversensing		
									A0712	Pacing Problem		
									A072201	High Impedance		
							E2403	No Clinical Signs, Symptoms or Conditions	A070909	Oversensing		
									A0712	Pacing Problem		
									A072201	High Impedance		
							F1001	Absence of Treatment	A070909	Oversensing		
									A0712	Pacing Problem		
									A072201	High Impedance		
							F15	Recognized Device or Procedural Complication	A070909	Oversensing		
									A0712	Pacing Problem		
									A072201	High Impedance		
							F2203	Imaging Required	A070909	Oversensing		
									A0712	Pacing Problem		
									A072201	High Impedance		
7105	2021	All other reportable incidents	35852	Single-chamber implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
7106	2021	All other reportable incidents	47265	Dual-chamber implantable pacemaker, rate-responsive	CT141	Electrodes/Leads and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT145	Cardiovascular devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		

					CT233	Surgical	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT2406	Implantable	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT2407	Active implantable	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT2727	Active implant devices/ancillaries	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT2750	Implantable cardiac electrical stimulators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT2985	Active	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT3001	Implantable metal	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT3002	Implantable synthetic polymer	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT3006	Implantable inorganic material	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT314	Electrically powered	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT321	Long-term surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT449	Generators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT981	Single use	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT983	Surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
7107	2021	All other reportable incidents	47265	Dual-chamber implantable pacemaker, rate-responsive	CT141	Electrodes/Leads and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	12/08/2021	12/08/2021
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	12/08/2021	12/08/2021
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	12/08/2021	12/08/2021
					CT145	Cardiovascular devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	12/08/2021	12/08/2021
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	12/08/2021	12/08/2021
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	12/08/2021	12/08/2021
					CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	12/08/2021	12/08/2021
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	12/08/2021	12/08/2021
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	12/08/2021	12/08/2021
					CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	12/08/2021	12/08/2021
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	12/08/2021	12/08/2021
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	12/08/2021	12/08/2021
					CT233	Surgical	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	12/08/2021	12/08/2021
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	12/08/2021	12/08/2021
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	12/08/2021	12/08/2021
					CT2406	Implantable	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	12/08/2021	12/08/2021
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	12/08/2021	12/08/2021
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	12/08/2021	12/08/2021
					CT2407	Active implantable	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	12/08/2021	12/08/2021
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	12/08/2021	12/08/2021
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	12/08/2021	12/08/2021
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	12/08/2021	12/08/2021
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	12/08/2021	12/08/2021
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	12/08/2021	12/08/2021
					CT2727	Active implant devices/ancillaries	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	12/08/2021	12/08/2021
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	12/08/2021	12/08/2021
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	12/08/2021	12/08/2021
					CT2750	Implantable cardiac electrical stimulators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	12/08/2021	12/08/2021
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	12/08/2021	12/08/2021
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	12/08/2021	12/08/2021
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	12/08/2021	12/08/2021
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	12/08/2021	12/08/2021
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	12/08/2021	12/08/2021
					CT2985	Active	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	12/08/2021	12/08/2021
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	12/08/2021	12/08/2021
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	12/08/2021	12/08/2021
					CT3001	Implantable metal	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	12/08/2021	12/08/2021
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	12/08/2021	12/08/2021
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	12/08/2021	12/08/2021
					CT3002	Implantable synthetic polymer	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	12/08/2021	12/08/2021
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	12/08/2021	12/08/2021
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	12/08/2021	12/08/2021
					CT3006	Implantable inorganic material	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	12/08/2021	12/08/2021
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	12/08/2021	12/08/2021
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	12/08/2021	12/08/2021
					CT314	Electrically powered	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	12/08/2021	12/08/2021
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	12/08/2021	12/08/2021
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	12/08/2021	12/08/2021
					CT321	Long-term surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	12/08/2021	12/08/2021
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	12/08/2021	12/08/2021
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	12/08/2021	12/08/2021
					CT449	Generators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	12/08/2021	12/08/2021
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	12/08/2021	12/08/2021
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	12/08/2021	12/08/2021
					CT981	Single use	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	12/08/2021	12/08/2021
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	12/08/2021	12/08/2021

						F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	12/08/2021	12/08/2021
				CT983	Surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	12/08/2021	12/08/2021
						F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	12/08/2021	12/08/2021
						F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	12/08/2021	12/08/2021
7108	2021	All other reportable incidents	47265	Dual-chamber implantable pacemaker, rate-responsive	CT141	Electrodes/Leads and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	21/10/2021	21/10/2021
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	21/10/2021	21/10/2021
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	21/10/2021	21/10/2021
					CT145	Cardiovascular devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	21/10/2021	21/10/2021
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	21/10/2021	21/10/2021
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	21/10/2021	21/10/2021
					CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	21/10/2021	21/10/2021
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	21/10/2021	21/10/2021
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	21/10/2021	21/10/2021
					CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	21/10/2021	21/10/2021
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	21/10/2021	21/10/2021
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	21/10/2021	21/10/2021
					CT233	Surgical	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	21/10/2021	21/10/2021
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	21/10/2021	21/10/2021
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	21/10/2021	21/10/2021
					CT2406	Implantable	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	21/10/2021	21/10/2021
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	21/10/2021	21/10/2021
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	21/10/2021	21/10/2021
					CT2407	Active implantable	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	21/10/2021	21/10/2021
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	21/10/2021	21/10/2021
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	21/10/2021	21/10/2021
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	21/10/2021	21/10/2021
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	21/10/2021	21/10/2021
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	21/10/2021	21/10/2021
					CT2727	Active implant devices/ancillaries	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	21/10/2021	21/10/2021
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	21/10/2021	21/10/2021
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	21/10/2021	21/10/2021
					CT2750	Implantable cardiac electrical stimulators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	21/10/2021	21/10/2021
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	21/10/2021	21/10/2021
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	21/10/2021	21/10/2021
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	21/10/2021	21/10/2021
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	21/10/2021	21/10/2021
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	21/10/2021	21/10/2021
					CT2985	Active	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	21/10/2021	21/10/2021
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	21/10/2021	21/10/2021
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	21/10/2021	21/10/2021
					CT3001	Implantable metal	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	21/10/2021	21/10/2021
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	21/10/2021	21/10/2021
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	21/10/2021	21/10/2021
					CT3002	Implantable synthetic polymer	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	21/10/2021	21/10/2021
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	21/10/2021	21/10/2021
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	21/10/2021	21/10/2021
					CT3006	Implantable inorganic material	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	21/10/2021	21/10/2021
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	21/10/2021	21/10/2021
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	21/10/2021	21/10/2021
					CT314	Electrically powered	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	21/10/2021	21/10/2021
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	21/10/2021	21/10/2021
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	21/10/2021	21/10/2021
					CT321	Long-term surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	21/10/2021	21/10/2021
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	21/10/2021	21/10/2021
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	21/10/2021	21/10/2021
					CT449	Generators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	21/10/2021	21/10/2021
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	21/10/2021	21/10/2021
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	21/10/2021	21/10/2021
					CT981	Single use	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	21/10/2021	21/10/2021
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	21/10/2021	21/10/2021
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	21/10/2021	21/10/2021
					CT983	Surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	21/10/2021	21/10/2021
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	21/10/2021	21/10/2021
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	21/10/2021	21/10/2021
7109	2021	All other reportable incidents	47265	Dual-chamber implantable pacemaker, rate-responsive	CT141	Electrodes/Leads and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT145	Cardiovascular devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT233	Surgical	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT2406	Implantable	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT2407	Active implantable	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT2727	Active implant devices/ancillaries	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT2750	Implantable cardiac electrical stimulators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT2985	Active	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		

					CT3001	Implantable metal	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT3002	Implantable synthetic polymer	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT3006	Implantable inorganic material	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT314	Electrically powered	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT321	Long-term surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT449	Generators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT981	Single use	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT983	Surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
7110	2021	All other reportable incidents	47263	Cardiac resynchronization therapy implantable pacemaker	CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	10/11/2021	10/11/2021
							F08	Hospitalization or Prolonged Hospitalization	A24	No Apparent Device Problem	10/11/2021	10/11/2021
							F1905	Device Revision or Replacement	A24	No Apparent Device Problem	10/11/2021	10/11/2021
					CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	10/11/2021	10/11/2021
							F08	Hospitalization or Prolonged Hospitalization	A24	No Apparent Device Problem	10/11/2021	10/11/2021
							F1905	Device Revision or Replacement	A24	No Apparent Device Problem	10/11/2021	10/11/2021
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	10/11/2021	10/11/2021
							F08	Hospitalization or Prolonged Hospitalization	A24	No Apparent Device Problem	10/11/2021	10/11/2021
							F1905	Device Revision or Replacement	A24	No Apparent Device Problem	10/11/2021	10/11/2021
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	10/11/2021	10/11/2021
							F08	Hospitalization or Prolonged Hospitalization	A24	No Apparent Device Problem	10/11/2021	10/11/2021
							F1905	Device Revision or Replacement	A24	No Apparent Device Problem	10/11/2021	10/11/2021
7111	2022	All other reportable incidents	47265	Dual-chamber implantable pacemaker, rate-responsive	CT141	Electrodes/Leads and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
					CT145	Cardiovascular devices	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
					CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
					CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
					CT233	Surgical	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
					CT2406	Implantable	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
					CT2407	Active implantable	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
					CT2727	Active implant devices/ancillaries	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
					CT2750	Implantable cardiac electrical stimulators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
					CT2985	Active	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
					CT3001	Implantable metal	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
					CT3002	Implantable synthetic polymer	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
					CT3006	Implantable inorganic material	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		

					CT314	Electrically powered	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
					CT321	Long-term surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
					CT449	Generators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
					CT981	Single use	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
					CT983	Surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
7112	2022	All other reportable incidents	17882	Physiologic-monitoring defibrillation system	CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070803	Failure to Power Up		
							F27	Problem Identified During Non-Clinical Procedure	A070803	Failure to Power Up		
7113	2022	All other reportable incidents	17882	Physiologic-monitoring defibrillation system	CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070803	Failure to Power Up		
							F27	Problem Identified During Non-Clinical Procedure	A070803	Failure to Power Up		
7114	2022	All other reportable incidents	17882	Physiologic-monitoring defibrillation system	CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A0719	Unexpected Shutdown		
							F26	No Health Consequences or Impact	A0719	Unexpected Shutdown		
7115	2021	All other reportable incidents	47265	Dual-chamber implantable pacemaker, rate-responsive	CT141	Electrodes/Leads and associated devices	E2104	Electric Shock	A070504	Premature Discharge of Battery		
							E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT145	Cardiovascular devices	E2104	Electric Shock	A070504	Premature Discharge of Battery		
							E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT1684	Cardiac pacemakers and associated devices	E2104	Electric Shock	A070504	Premature Discharge of Battery		
							E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT1685	Cardiac pacemakers	E2104	Electric Shock	A070504	Premature Discharge of Battery		
							E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT233	Surgical	E2104	Electric Shock	A070504	Premature Discharge of Battery		
							E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT2406	Implantable	E2104	Electric Shock	A070504	Premature Discharge of Battery		
							E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT2407	Active implantable	E2104	Electric Shock	A070504	Premature Discharge of Battery		
							E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT257	Implantable cardiac pacemakers	E2104	Electric Shock	A070504	Premature Discharge of Battery		
							E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT2727	Active implant devices/ancillaries	E2104	Electric Shock	A070504	Premature Discharge of Battery		
							E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT2750	Implantable cardiac electrical stimulators and associated devices	E2104	Electric Shock	A070504	Premature Discharge of Battery		
							E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT2866	Implantable cardiac electrical stimulators	E2104	Electric Shock	A070504	Premature Discharge of Battery		
							E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT2985	Active	E2104	Electric Shock	A070504	Premature Discharge of Battery		
							E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT3001	Implantable metal	E2104	Electric Shock	A070504	Premature Discharge of Battery		
							E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT3002	Implantable synthetic polymer	E2104	Electric Shock	A070504	Premature Discharge of Battery		
							E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT3006	Implantable inorganic material	E2104	Electric Shock	A070504	Premature Discharge of Battery		
							E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT314	Electrically powered	E2104	Electric Shock	A070504	Premature Discharge of Battery		
							E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT321	Long-term surgical invasive	E2104	Electric Shock	A070504	Premature Discharge of Battery		
							E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT449	Generators and associated devices	E2104	Electric Shock	A070504	Premature Discharge of Battery		
							E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT981	Single use	E2104	Electric Shock	A070504	Premature Discharge of Battery		
							E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT983	Surgical invasive	E2104	Electric Shock	A070504	Premature Discharge of Battery		
							E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
7116	2021	All other reportable incidents	36241	Pacemaker/defibrillator lead	CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A040101	Fracture		
									A070101	Failure to Capture		

									A072201	High Impedance		
							F1905	Device Revision or Replacement	A040101	Fracture		
									A070101	Failure to Capture		
									A072201	High Impedance		
7117	2022	All other reportable incidents	47205	Cardiac pulse generator programmer	CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A0902	Display or Visual Feedback Problem		
							F26	No Health Consequences or Impact	A0902	Display or Visual Feedback Problem		
7118	2022	All other reportable incidents	36241	Pacemaker/defibrillator lead	CT1684	Cardiac pacemakers and associated devices	E2104	Electric Shock	A070101	Failure to Capture		
									A070909	Oversensing		
									A0712	Pacing Problem		
									A071302	Inappropriate/Inadequate Shock/Stimulation		
									A072201	High Impedance		
							E2403	No Clinical Signs, Symptoms or Conditions	A070101	Failure to Capture		
									A070909	Oversensing		
									A0712	Pacing Problem		
									A071302	Inappropriate/Inadequate Shock/Stimulation		
									A072201	High Impedance		
							F08	Hospitalization or Prolonged Hospitalization	A070101	Failure to Capture		
									A070909	Oversensing		
									A0712	Pacing Problem		
									A071302	Inappropriate/Inadequate Shock/Stimulation		
									A072201	High Impedance		
							F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A070101	Failure to Capture		
									A070909	Oversensing		
									A0712	Pacing Problem		
									A071302	Inappropriate/Inadequate Shock/Stimulation		
									A072201	High Impedance		
							F1905	Device Revision or Replacement	A070101	Failure to Capture		
									A070909	Oversensing		
									A0712	Pacing Problem		
									A071302	Inappropriate/Inadequate Shock/Stimulation		
									A072201	High Impedance		
7119	2021	All other reportable incidents	35852	Single-chamber implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem	07/12/2021	07/12/2021
									A070504	Premature Discharge of Battery	07/12/2021	07/12/2021
							F08	Hospitalization or Prolonged Hospitalization	A0705	Battery Problem	07/12/2021	07/12/2021
									A070504	Premature Discharge of Battery	07/12/2021	07/12/2021
							F1905	Device Revision or Replacement	A0705	Battery Problem	07/12/2021	07/12/2021
									A070504	Premature Discharge of Battery	07/12/2021	07/12/2021
7120	2021	All other reportable incidents	37265	Dual-chamber implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	27/10/2021	27/10/2021
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	27/10/2021	27/10/2021
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	27/10/2021	27/10/2021
7121	2022	All other reportable incidents	36241	Pacemaker/defibrillator lead	CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A040101	Fracture		
									A070909	Oversensing		
							F08	Hospitalization or Prolonged Hospitalization	A040101	Fracture		
									A070909	Oversensing		
							F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A040101	Fracture		
									A070909	Oversensing		
							F23	Unexpected Medical Intervention	A040101	Fracture		
									A070909	Oversensing		
7122	2021	All other reportable incidents	47265	Dual-chamber implantable pacemaker, rate-responsive	CT141	Electrodes/Leads and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem	31/01/2022	31/01/2022
									A070504	Premature Discharge of Battery	31/01/2022	31/01/2022
							F08	Hospitalization or Prolonged Hospitalization	A0705	Battery Problem	31/01/2022	31/01/2022
									A070504	Premature Discharge of Battery	31/01/2022	31/01/2022
							F1903	Device Explantation	A0705	Battery Problem	31/01/2022	31/01/2022
									A070504	Premature Discharge of Battery	31/01/2022	31/01/2022
					CT145	Cardiovascular devices	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem	31/01/2022	31/01/2022
									A070504	Premature Discharge of Battery	31/01/2022	31/01/2022
							F08	Hospitalization or Prolonged Hospitalization	A0705	Battery Problem	31/01/2022	31/01/2022
									A070504	Premature Discharge of Battery	31/01/2022	31/01/2022
							F1903	Device Explantation	A0705	Battery Problem	31/01/2022	31/01/2022
									A070504	Premature Discharge of Battery	31/01/2022	31/01/2022
					CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem	31/01/2022	31/01/2022
									A070504	Premature Discharge of Battery	31/01/2022	31/01/2022
							F08	Hospitalization or Prolonged Hospitalization	A0705	Battery Problem	31/01/2022	31/01/2022
									A070504	Premature Discharge of Battery	31/01/2022	31/01/2022
							F1903	Device Explantation	A0705	Battery Problem	31/01/2022	31/01/2022
									A070504	Premature Discharge of Battery	31/01/2022	31/01/2022
					CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem	31/01/2022	31/01/2022
									A070504	Premature Discharge of Battery	31/01/2022	31/01/2022
							F08	Hospitalization or Prolonged Hospitalization	A0705	Battery Problem	31/01/2022	31/01/2022
									A070504	Premature Discharge of Battery	31/01/2022	31/01/2022
							F1903	Device Explantation	A0705	Battery Problem	31/01/2022	31/01/2022
									A070504	Premature Discharge of Battery	31/01/2022	31/01/2022
					CT233	Surgical	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem	31/01/2022	31/01/2022
									A070504	Premature Discharge of Battery	31/01/2022	31/01/2022
							F08	Hospitalization or Prolonged Hospitalization	A0705	Battery Problem	31/01/2022	31/01/2022
									A070504	Premature Discharge of Battery	31/01/2022	31/01/2022
							F1903	Device Explantation	A0705	Battery Problem	31/01/2022	31/01/2022
									A070504	Premature Discharge of Battery	31/01/2022	31/01/2022
					CT2406	Implantable	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem	31/01/2022	31/01/2022
									A070504	Premature Discharge of Battery	31/01/2022	31/01/2022
							F08	Hospitalization or Prolonged Hospitalization	A0705	Battery Problem	31/01/2022	31/01/2022
									A070504	Premature Discharge of Battery	31/01/2022	31/01/2022
							F1903	Device Explantation	A0705	Battery Problem	31/01/2022	31/01/2022
									A070504	Premature Discharge of Battery	31/01/2022	31/01/2022
					CT2407	Active implantable	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem	31/01/2022	31/01/2022
									A070504	Premature Discharge of Battery	31/01/2022	31/01/2022
							F08	Hospitalization or Prolonged Hospitalization	A0705	Battery Problem	31/01/2022	31/01/2022
									A070504	Premature Discharge of Battery	31/01/2022	31/01/2022

							F1903	Device Explantation	A0705	Battery Problem	31/01/2022	31/01/2022
									A070504	Premature Discharge of Battery	31/01/2022	31/01/2022
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem	31/01/2022	31/01/2022
									A070504	Premature Discharge of Battery	31/01/2022	31/01/2022
							F08	Hospitalization or Prolonged Hospitalization	A0705	Battery Problem	31/01/2022	31/01/2022
									A070504	Premature Discharge of Battery	31/01/2022	31/01/2022
							F1903	Device Explantation	A0705	Battery Problem	31/01/2022	31/01/2022
									A070504	Premature Discharge of Battery	31/01/2022	31/01/2022
					CT2727	Active implant devices/ancillaries	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem	31/01/2022	31/01/2022
									A070504	Premature Discharge of Battery	31/01/2022	31/01/2022
							F08	Hospitalization or Prolonged Hospitalization	A0705	Battery Problem	31/01/2022	31/01/2022
									A070504	Premature Discharge of Battery	31/01/2022	31/01/2022
							F1903	Device Explantation	A0705	Battery Problem	31/01/2022	31/01/2022
									A070504	Premature Discharge of Battery	31/01/2022	31/01/2022
					CT2750	Implantable cardiac electrical stimulators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem	31/01/2022	31/01/2022
									A070504	Premature Discharge of Battery	31/01/2022	31/01/2022
							F08	Hospitalization or Prolonged Hospitalization	A0705	Battery Problem	31/01/2022	31/01/2022
									A070504	Premature Discharge of Battery	31/01/2022	31/01/2022
							F1903	Device Explantation	A0705	Battery Problem	31/01/2022	31/01/2022
									A070504	Premature Discharge of Battery	31/01/2022	31/01/2022
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem	31/01/2022	31/01/2022
									A070504	Premature Discharge of Battery	31/01/2022	31/01/2022
							F08	Hospitalization or Prolonged Hospitalization	A0705	Battery Problem	31/01/2022	31/01/2022
									A070504	Premature Discharge of Battery	31/01/2022	31/01/2022
							F1903	Device Explantation	A0705	Battery Problem	31/01/2022	31/01/2022
									A070504	Premature Discharge of Battery	31/01/2022	31/01/2022
					CT2985	Active	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem	31/01/2022	31/01/2022
									A070504	Premature Discharge of Battery	31/01/2022	31/01/2022
							F08	Hospitalization or Prolonged Hospitalization	A0705	Battery Problem	31/01/2022	31/01/2022
									A070504	Premature Discharge of Battery	31/01/2022	31/01/2022
							F1903	Device Explantation	A0705	Battery Problem	31/01/2022	31/01/2022
									A070504	Premature Discharge of Battery	31/01/2022	31/01/2022
					CT3001	Implantable metal	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem	31/01/2022	31/01/2022
									A070504	Premature Discharge of Battery	31/01/2022	31/01/2022
							F08	Hospitalization or Prolonged Hospitalization	A0705	Battery Problem	31/01/2022	31/01/2022
									A070504	Premature Discharge of Battery	31/01/2022	31/01/2022
							F1903	Device Explantation	A0705	Battery Problem	31/01/2022	31/01/2022
									A070504	Premature Discharge of Battery	31/01/2022	31/01/2022
					CT3002	Implantable synthetic polymer	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem	31/01/2022	31/01/2022
									A070504	Premature Discharge of Battery	31/01/2022	31/01/2022
							F08	Hospitalization or Prolonged Hospitalization	A0705	Battery Problem	31/01/2022	31/01/2022
									A070504	Premature Discharge of Battery	31/01/2022	31/01/2022
							F1903	Device Explantation	A0705	Battery Problem	31/01/2022	31/01/2022
									A070504	Premature Discharge of Battery	31/01/2022	31/01/2022
					CT3006	Implantable inorganic material	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem	31/01/2022	31/01/2022
									A070504	Premature Discharge of Battery	31/01/2022	31/01/2022
							F08	Hospitalization or Prolonged Hospitalization	A0705	Battery Problem	31/01/2022	31/01/2022
									A070504	Premature Discharge of Battery	31/01/2022	31/01/2022
							F1903	Device Explantation	A0705	Battery Problem	31/01/2022	31/01/2022
									A070504	Premature Discharge of Battery	31/01/2022	31/01/2022
					CT314	Electrically powered	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem	31/01/2022	31/01/2022
									A070504	Premature Discharge of Battery	31/01/2022	31/01/2022
							F08	Hospitalization or Prolonged Hospitalization	A0705	Battery Problem	31/01/2022	31/01/2022
									A070504	Premature Discharge of Battery	31/01/2022	31/01/2022
							F1903	Device Explantation	A0705	Battery Problem	31/01/2022	31/01/2022
									A070504	Premature Discharge of Battery	31/01/2022	31/01/2022
					CT321	Long-term surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem	31/01/2022	31/01/2022
									A070504	Premature Discharge of Battery	31/01/2022	31/01/2022
							F08	Hospitalization or Prolonged Hospitalization	A0705	Battery Problem	31/01/2022	31/01/2022
									A070504	Premature Discharge of Battery	31/01/2022	31/01/2022
							F1903	Device Explantation	A0705	Battery Problem	31/01/2022	31/01/2022
									A070504	Premature Discharge of Battery	31/01/2022	31/01/2022
					CT449	Generators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem	31/01/2022	31/01/2022
									A070504	Premature Discharge of Battery	31/01/2022	31/01/2022
							F08	Hospitalization or Prolonged Hospitalization	A0705	Battery Problem	31/01/2022	31/01/2022
									A070504	Premature Discharge of Battery	31/01/2022	31/01/2022
							F1903	Device Explantation	A0705	Battery Problem	31/01/2022	31/01/2022
									A070504	Premature Discharge of Battery	31/01/2022	31/01/2022
					CT981	Single use	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem	31/01/2022	31/01/2022
									A070504	Premature Discharge of Battery	31/01/2022	31/01/2022
							F08	Hospitalization or Prolonged Hospitalization	A0705	Battery Problem	31/01/2022	31/01/2022
									A070504	Premature Discharge of Battery	31/01/2022	31/01/2022
							F1903	Device Explantation	A0705	Battery Problem	31/01/2022	31/01/2022
									A070504	Premature Discharge of Battery	31/01/2022	31/01/2022
					CT983	Surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem	31/01/2022	31/01/2022
									A070504	Premature Discharge of Battery	31/01/2022	31/01/2022
							F08	Hospitalization or Prolonged Hospitalization	A0705	Battery Problem	31/01/2022	31/01/2022
									A070504	Premature Discharge of Battery	31/01/2022	31/01/2022
							F1903	Device Explantation	A0705	Battery Problem	31/01/2022	31/01/2022
									A070504	Premature Discharge of Battery	31/01/2022	31/01/2022
7123	2022	All other reportable incidents	47263	Cardiac resynchronization therapy implantable pacemaker	CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A0908	Incorrect, Inadequate or Imprecise Result or Readings	19/07/2022	19/07/2022
									A160503	Inappropriate or Unexpected Reset	19/07/2022	19/07/2022
							F08	Hospitalization or Prolonged Hospitalization	A0908	Incorrect, Inadequate or Imprecise Result or Readings	19/07/2022	19/07/2022
									A160503	Inappropriate or Unexpected Reset	19/07/2022	19/07/2022
							F1903	Device Explantation	A0908	Incorrect, Inadequate or Imprecise Result or Readings	19/07/2022	19/07/2022

								A160503	Inappropriate or Unexpected Reset	19/07/2022	19/07/2022
				CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A0908	Incorrect, Inadequate or Imprecise Result or Readings	19/07/2022	19/07/2022
								A160503	Inappropriate or Unexpected Reset	19/07/2022	19/07/2022
						F08	Hospitalization or Prolonged Hospitalization	A0908	Incorrect, Inadequate or Imprecise Result or Readings	19/07/2022	19/07/2022
								A160503	Inappropriate or Unexpected Reset	19/07/2022	19/07/2022
						F1903	Device Explantation	A0908	Incorrect, Inadequate or Imprecise Result or Readings	19/07/2022	19/07/2022
								A160503	Inappropriate or Unexpected Reset	19/07/2022	19/07/2022
				CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A0908	Incorrect, Inadequate or Imprecise Result or Readings	19/07/2022	19/07/2022
								A160503	Inappropriate or Unexpected Reset	19/07/2022	19/07/2022
						F08	Hospitalization or Prolonged Hospitalization	A0908	Incorrect, Inadequate or Imprecise Result or Readings	19/07/2022	19/07/2022
								A160503	Inappropriate or Unexpected Reset	19/07/2022	19/07/2022
						F1903	Device Explantation	A0908	Incorrect, Inadequate or Imprecise Result or Readings	19/07/2022	19/07/2022
								A160503	Inappropriate or Unexpected Reset	19/07/2022	19/07/2022
				CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A0908	Incorrect, Inadequate or Imprecise Result or Readings	19/07/2022	19/07/2022
								A160503	Inappropriate or Unexpected Reset	19/07/2022	19/07/2022
						F08	Hospitalization or Prolonged Hospitalization	A0908	Incorrect, Inadequate or Imprecise Result or Readings	19/07/2022	19/07/2022
								A160503	Inappropriate or Unexpected Reset	19/07/2022	19/07/2022
						F1903	Device Explantation	A0908	Incorrect, Inadequate or Imprecise Result or Readings	19/07/2022	19/07/2022
								A160503	Inappropriate or Unexpected Reset	19/07/2022	19/07/2022
7124	2021	All other reportable incidents	47265	Dual-chamber implantable pacemaker, rate-responsive	CT141	Electrodes/Leads and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	30/12/2021	30/12/2021
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	30/12/2021	30/12/2021
							F1903	Device Explantation	A070504	Premature Discharge of Battery	30/12/2021	30/12/2021
					CT145	Cardiovascular devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	30/12/2021	30/12/2021
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	30/12/2021	30/12/2021
							F1903	Device Explantation	A070504	Premature Discharge of Battery	30/12/2021	30/12/2021
					CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	30/12/2021	30/12/2021
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	30/12/2021	30/12/2021
							F1903	Device Explantation	A070504	Premature Discharge of Battery	30/12/2021	30/12/2021
					CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	30/12/2021	30/12/2021
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	30/12/2021	30/12/2021
							F1903	Device Explantation	A070504	Premature Discharge of Battery	30/12/2021	30/12/2021
					CT233	Surgical	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	30/12/2021	30/12/2021
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	30/12/2021	30/12/2021
							F1903	Device Explantation	A070504	Premature Discharge of Battery	30/12/2021	30/12/2021
					CT2406	Implantable	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	30/12/2021	30/12/2021
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	30/12/2021	30/12/2021
							F1903	Device Explantation	A070504	Premature Discharge of Battery	30/12/2021	30/12/2021
					CT2407	Active implantable	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	30/12/2021	30/12/2021
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	30/12/2021	30/12/2021
							F1903	Device Explantation	A070504	Premature Discharge of Battery	30/12/2021	30/12/2021
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	30/12/2021	30/12/2021
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	30/12/2021	30/12/2021
							F1903	Device Explantation	A070504	Premature Discharge of Battery	30/12/2021	30/12/2021
					CT2727	Active implant devices/ancillaries	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	30/12/2021	30/12/2021
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	30/12/2021	30/12/2021
							F1903	Device Explantation	A070504	Premature Discharge of Battery	30/12/2021	30/12/2021
					CT2750	Implantable cardiac electrical stimulators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	30/12/2021	30/12/2021
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	30/12/2021	30/12/2021
							F1903	Device Explantation	A070504	Premature Discharge of Battery	30/12/2021	30/12/2021
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	30/12/2021	30/12/2021
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	30/12/2021	30/12/2021
							F1903	Device Explantation	A070504	Premature Discharge of Battery	30/12/2021	30/12/2021
					CT2985	Active	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	30/12/2021	30/12/2021
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	30/12/2021	30/12/2021
							F1903	Device Explantation	A070504	Premature Discharge of Battery	30/12/2021	30/12/2021
					CT3001	Implantable metal	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	30/12/2021	30/12/2021
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	30/12/2021	30/12/2021
							F1903	Device Explantation	A070504	Premature Discharge of Battery	30/12/2021	30/12/2021
					CT3002	Implantable synthetic polymer	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	30/12/2021	30/12/2021
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	30/12/2021	30/12/2021
							F1903	Device Explantation	A070504	Premature Discharge of Battery	30/12/2021	30/12/2021
					CT3006	Implantable inorganic material	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	30/12/2021	30/12/2021
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	30/12/2021	30/12/2021
							F1903	Device Explantation	A070504	Premature Discharge of Battery	30/12/2021	30/12/2021
					CT314	Electrically powered	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	30/12/2021	30/12/2021
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	30/12/2021	30/12/2021
							F1903	Device Explantation	A070504	Premature Discharge of Battery	30/12/2021	30/12/2021
					CT321	Long-term surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	30/12/2021	30/12/2021
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	30/12/2021	30/12/2021
							F1903	Device Explantation	A070504	Premature Discharge of Battery	30/12/2021	30/12/2021
					CT449	Generators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	30/12/2021	30/12/2021
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	30/12/2021	30/12/2021
							F1903	Device Explantation	A070504	Premature Discharge of Battery	30/12/2021	30/12/2021
					CT981	Single use	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	30/12/2021	30/12/2021
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	30/12/2021	30/12/2021
							F1903	Device Explantation	A070504	Premature Discharge of Battery	30/12/2021	30/12/2021
					CT983	Surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	30/12/2021	30/12/2021
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	30/12/2021	30/12/2021
							F1903	Device Explantation	A070504	Premature Discharge of Battery	30/12/2021	30/12/2021
7125	2022	All other reportable incidents	35852	Single-chamber implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E2104	Electric Shock	A070504	Premature Discharge of Battery	25/11/2022	25/11/2022
									A071302	Inappropriate/Inadequate Shock/Stimulation	25/11/2022	25/11/2022

									A090801	Signal Artifact/Noise	25/11/2022	25/11/2022
							E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	25/11/2022	25/11/2022
									A071302	Inappropriate/Inadequate Shock/Stimulation	25/11/2022	25/11/2022
									A090801	Signal Artifact/Noise	25/11/2022	25/11/2022
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	25/11/2022	25/11/2022
									A071302	Inappropriate/Inadequate Shock/Stimulation	25/11/2022	25/11/2022
									A090801	Signal Artifact/Noise	25/11/2022	25/11/2022
							F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A070504	Premature Discharge of Battery	25/11/2022	25/11/2022
									A071302	Inappropriate/Inadequate Shock/Stimulation	25/11/2022	25/11/2022
									A090801	Signal Artifact/Noise	25/11/2022	25/11/2022
							F1903	Device Explantation	A070504	Premature Discharge of Battery	25/11/2022	25/11/2022
									A071302	Inappropriate/Inadequate Shock/Stimulation	25/11/2022	25/11/2022
									A090801	Signal Artifact/Noise	25/11/2022	25/11/2022
							F2203	Imaging Required	A070504	Premature Discharge of Battery	25/11/2022	25/11/2022
									A071302	Inappropriate/Inadequate Shock/Stimulation	25/11/2022	25/11/2022
									A090801	Signal Artifact/Noise	25/11/2022	25/11/2022
7126	2022	All other reportable incidents	35223	Endocardial/interventricular septal pacing lead	CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070909	Oversensing		
							F26	No Health Consequences or Impact	A070909	Oversensing		
7127	2022	All other reportable incidents	35223	Endocardial/interventricular septal pacing lead	CT1684	Cardiac pacemakers and associated devices	E2402	Appropriate Clinical Signs, Symptoms and Conditions Term/Code not Available	A05	Mechanical Problem		
									A070101	Failure to Capture		
									A070909	Oversensing		
									A090801	Signal Artifact/Noise		
							E2403	No Clinical Signs, Symptoms or Conditions	A05	Mechanical Problem		
									A070101	Failure to Capture		
									A070909	Oversensing		
									A090801	Signal Artifact/Noise		
							F08	Hospitalization or Prolonged Hospitalization	A05	Mechanical Problem		
									A070101	Failure to Capture		
									A070909	Oversensing		
									A090801	Signal Artifact/Noise		
							F12	Serious Injury/ Illness/ Impairment	A05	Mechanical Problem		
									A070101	Failure to Capture		
									A070909	Oversensing		
									A090801	Signal Artifact/Noise		
							F23	Unexpected Medical Intervention	A05	Mechanical Problem		
									A070101	Failure to Capture		
									A070909	Oversensing		
									A090801	Signal Artifact/Noise		
7128	2022	All other reportable incidents	36241	Pacemaker/defibrillator lead	CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A0203	Defective Device	30/06/2022	30/06/2022
							F1908	Prolonged Surgery	A0203	Defective Device	30/06/2022	30/06/2022
7129	2022	All other reportable incidents	47205	Cardiac pulse generator programmer	CT1684	Cardiac pacemakers and associated devices	E0602	Cardiac Arrest	A0712	Pacing Problem		
									A1104	Computer Operating System Problem		
							E2403	No Clinical Signs, Symptoms or Conditions	A0712	Pacing Problem		
									A1104	Computer Operating System Problem		
							F1203	Life Threatening Illness or Injury	A0712	Pacing Problem		
									A1104	Computer Operating System Problem		
7130	2021	All other reportable incidents	63188	Left-ventricular wireless pacing system sensor/transmitter	CT141	Electrodes/Leads and associated devices			A070801	Complete Loss of Power		
					CT145	Cardiovascular devices			A070801	Complete Loss of Power		
					CT1577	Wireless monitoring system transmitters			A070801	Complete Loss of Power		
					CT1684	Cardiac pacemakers and associated devices			A070801	Complete Loss of Power		
					CT2317	Transmitters/Receivers/Transceivers			A070801	Complete Loss of Power		
					CT233	Surgical			A070801	Complete Loss of Power		
					CT2406	Implantable			A070801	Complete Loss of Power		
					CT2407	Active implantable			A070801	Complete Loss of Power		
					CT255	Transducers and associated devices			A070801	Complete Loss of Power		
					CT2726	Physiological monitoring/measuring devices			A070801	Complete Loss of Power		
					CT2727	Active implant devices/ancillaries			A070801	Complete Loss of Power		
					CT2750	Implantable cardiac electrical stimulators and associated devices			A070801	Complete Loss of Power		
					CT2885	Physiological monitoring/measuring system transmitters/receivers/transceivers			A070801	Complete Loss of Power		
					CT2985	Active			A070801	Complete Loss of Power		
					CT3001	Implantable metal			A070801	Complete Loss of Power		
					CT3002	Implantable synthetic polymer			A070801	Complete Loss of Power		
					CT3006	Implantable inorganic material			A070801	Complete Loss of Power		
					CT314	Electrically powered			A070801	Complete Loss of Power		
					CT321	Long-term surgical invasive			A070801	Complete Loss of Power		
					CT384	Sensors and associated devices			A070801	Complete Loss of Power		
					CT449	Generators and associated devices			A070801	Complete Loss of Power		
					CT981	Single use			A070801	Complete Loss of Power		
					CT983	Surgical invasive			A070801	Complete Loss of Power		
					CT994	Sensors			A070801	Complete Loss of Power		
7131	2021	All other reportable incidents	17882	Physiologic-monitoring defibrillation system	CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A090101	Inappropriate Audible Prompt/Feedback		
							F27	Problem Identified During Non-Clinical Procedure	A090101	Inappropriate Audible Prompt/Feedback		
7132	2021	All other reportable incidents	47263	Cardiac resynchronization therapy implantable pacemaker	CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A1304	Telemetry Discrepancy	05/11/2021	05/11/2021
							F08	Hospitalization or Prolonged Hospitalization	A1304	Telemetry Discrepancy	05/11/2021	05/11/2021
							F1905	Device Revision or Replacement	A1304	Telemetry Discrepancy	05/11/2021	05/11/2021
					CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A1304	Telemetry Discrepancy	05/11/2021	05/11/2021
							F08	Hospitalization or Prolonged Hospitalization	A1304	Telemetry Discrepancy	05/11/2021	05/11/2021
							F1905	Device Revision or Replacement	A1304	Telemetry Discrepancy	05/11/2021	05/11/2021
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A1304	Telemetry Discrepancy	05/11/2021	05/11/2021
							F08	Hospitalization or Prolonged Hospitalization	A1304	Telemetry Discrepancy	05/11/2021	05/11/2021
							F1905	Device Revision or Replacement	A1304	Telemetry Discrepancy	05/11/2021	05/11/2021
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A1304	Telemetry Discrepancy	05/11/2021	05/11/2021
							F08	Hospitalization or Prolonged Hospitalization	A1304	Telemetry Discrepancy	05/11/2021	05/11/2021
							F1905	Device Revision or Replacement	A1304	Telemetry Discrepancy	05/11/2021	05/11/2021
7133	2021	All other reportable incidents	47265	Dual-chamber implantable pacemaker, rate-responsive	CT141	Electrodes/Leads and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		

					CT145	Cardiovascular devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT233	Surgical	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT2406	Implantable	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT2407	Active implantable	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT2727	Active implant devices/ancillaries	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT2750	Implantable cardiac electrical stimulators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT2985	Active	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT3001	Implantable metal	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT3002	Implantable synthetic polymer	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT3006	Implantable inorganic material	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT314	Electrically powered	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT321	Long-term surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT449	Generators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT981	Single use	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT983	Surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
7134	2021	All other reportable incidents	35223	Endocardial/interventricular septal pacing lead	CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070102	High Capture Threshold		
									A070909	Oversensing		
									A072201	High Impedance		
							F26	No Health Consequences or Impact	A070102	High Capture Threshold		
									A070909	Oversensing		
									A072201	High Impedance		
7135	2021	All other reportable incidents	47263	Cardiac resynchronization therapy implantable pacemaker	CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	27/01/2022	27/01/2022
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	27/01/2022	27/01/2022
					CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	27/01/2022	27/01/2022
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	27/01/2022	27/01/2022
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	27/01/2022	27/01/2022
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	27/01/2022	27/01/2022
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	27/01/2022	27/01/2022
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	27/01/2022	27/01/2022
7136	2022	All other reportable incidents	60789	Intracardiac pacemaker	CT130	Catheters and associated devices	E011903	Syncope/Fainting	A071102	Failure to Interrogate		
									A071204	No Pacing		
							E2403	No Clinical Signs, Symptoms or Conditions	A071102	Failure to Interrogate		
									A071204	No Pacing		
							F1901	Additional Surgery	A071102	Failure to Interrogate		
									A071204	No Pacing		
					CT145	Cardiovascular devices	E011903	Syncope/Fainting	A071102	Failure to Interrogate		
									A071204	No Pacing		
							E2403	No Clinical Signs, Symptoms or Conditions	A071102	Failure to Interrogate		
									A071204	No Pacing		
							F1901	Additional Surgery	A071102	Failure to Interrogate		
									A071204	No Pacing		
					CT1684	Cardiac pacemakers and associated devices	E011903	Syncope/Fainting	A071102	Failure to Interrogate		
									A071204	No Pacing		
							E2403	No Clinical Signs, Symptoms or Conditions	A071102	Failure to Interrogate		
									A071204	No Pacing		
							F1901	Additional Surgery	A071102	Failure to Interrogate		
									A071204	No Pacing		
					CT1685	Cardiac pacemakers	E011903	Syncope/Fainting	A071102	Failure to Interrogate		
									A071204	No Pacing		
							E2403	No Clinical Signs, Symptoms or Conditions	A071102	Failure to Interrogate		
									A071204	No Pacing		
							F1901	Additional Surgery	A071102	Failure to Interrogate		
									A071204	No Pacing		
					CT233	Surgical	E011903	Syncope/Fainting	A071102	Failure to Interrogate		
									A071204	No Pacing		
							E2403	No Clinical Signs, Symptoms or Conditions	A071102	Failure to Interrogate		
									A071204	No Pacing		
							F1901	Additional Surgery	A071102	Failure to Interrogate		
									A071204	No Pacing		
					CT2406	Implantable	E011903	Syncope/Fainting	A071102	Failure to Interrogate		
									A071204	No Pacing		
							E2403	No Clinical Signs, Symptoms or Conditions	A071102	Failure to Interrogate		
									A071204	No Pacing		
							F1901	Additional Surgery	A071102	Failure to Interrogate		
									A071204	No Pacing		

					CT2407	Active implantable	E011903	Syncope/Fainting	A071102	Failure to Interrogate		
									A071204	No Pacing		
							E2403	No Clinical Signs, Symptoms or Conditions	A071102	Failure to Interrogate		
									A071204	No Pacing		
							F1901	Additional Surgery	A071102	Failure to Interrogate		
									A071204	No Pacing		
					CT257	Implantable cardiac pacemakers	E011903	Syncope/Fainting	A071102	Failure to Interrogate		
									A071204	No Pacing		
							E2403	No Clinical Signs, Symptoms or Conditions	A071102	Failure to Interrogate		
									A071204	No Pacing		
							F1901	Additional Surgery	A071102	Failure to Interrogate		
									A071204	No Pacing		
					CT2727	Active implant devices/ancillaries	E011903	Syncope/Fainting	A071102	Failure to Interrogate		
									A071204	No Pacing		
							E2403	No Clinical Signs, Symptoms or Conditions	A071102	Failure to Interrogate		
									A071204	No Pacing		
							F1901	Additional Surgery	A071102	Failure to Interrogate		
									A071204	No Pacing		
					CT2750	Implantable cardiac electrical stimulators and associated devices	E011903	Syncope/Fainting	A071102	Failure to Interrogate		
									A071204	No Pacing		
							E2403	No Clinical Signs, Symptoms or Conditions	A071102	Failure to Interrogate		
									A071204	No Pacing		
							F1901	Additional Surgery	A071102	Failure to Interrogate		
									A071204	No Pacing		
					CT2818	Catheterization/Cannulation/Fluid path devices	E011903	Syncope/Fainting	A071102	Failure to Interrogate		
									A071204	No Pacing		
							E2403	No Clinical Signs, Symptoms or Conditions	A071102	Failure to Interrogate		
									A071204	No Pacing		
							F1901	Additional Surgery	A071102	Failure to Interrogate		
									A071204	No Pacing		
					CT2866	Implantable cardiac electrical stimulators	E011903	Syncope/Fainting	A071102	Failure to Interrogate		
									A071204	No Pacing		
							E2403	No Clinical Signs, Symptoms or Conditions	A071102	Failure to Interrogate		
									A071204	No Pacing		
							F1901	Additional Surgery	A071102	Failure to Interrogate		
									A071204	No Pacing		
					CT2985	Active	E011903	Syncope/Fainting	A071102	Failure to Interrogate		
									A071204	No Pacing		
							E2403	No Clinical Signs, Symptoms or Conditions	A071102	Failure to Interrogate		
									A071204	No Pacing		
							F1901	Additional Surgery	A071102	Failure to Interrogate		
									A071204	No Pacing		
					CT3001	Implantable metal	E011903	Syncope/Fainting	A071102	Failure to Interrogate		
									A071204	No Pacing		
							E2403	No Clinical Signs, Symptoms or Conditions	A071102	Failure to Interrogate		
									A071204	No Pacing		
							F1901	Additional Surgery	A071102	Failure to Interrogate		
									A071204	No Pacing		
					CT3002	Implantable synthetic polymer	E011903	Syncope/Fainting	A071102	Failure to Interrogate		
									A071204	No Pacing		
							E2403	No Clinical Signs, Symptoms or Conditions	A071102	Failure to Interrogate		
									A071204	No Pacing		
							F1901	Additional Surgery	A071102	Failure to Interrogate		
									A071204	No Pacing		
					CT3006	Implantable inorganic material	E011903	Syncope/Fainting	A071102	Failure to Interrogate		
									A071204	No Pacing		
							E2403	No Clinical Signs, Symptoms or Conditions	A071102	Failure to Interrogate		
									A071204	No Pacing		
							F1901	Additional Surgery	A071102	Failure to Interrogate		
									A071204	No Pacing		
					CT314	Electrically powered	E011903	Syncope/Fainting	A071102	Failure to Interrogate		
									A071204	No Pacing		
							E2403	No Clinical Signs, Symptoms or Conditions	A071102	Failure to Interrogate		
									A071204	No Pacing		
							F1901	Additional Surgery	A071102	Failure to Interrogate		
									A071204	No Pacing		
					CT321	Long-term surgical invasive	E011903	Syncope/Fainting	A071102	Failure to Interrogate		
									A071204	No Pacing		
							E2403	No Clinical Signs, Symptoms or Conditions	A071102	Failure to Interrogate		
									A071204	No Pacing		
							F1901	Additional Surgery	A071102	Failure to Interrogate		
									A071204	No Pacing		
					CT981	Single use	E011903	Syncope/Fainting	A071102	Failure to Interrogate		
									A071204	No Pacing		
							E2403	No Clinical Signs, Symptoms or Conditions	A071102	Failure to Interrogate		
									A071204	No Pacing		
							F1901	Additional Surgery	A071102	Failure to Interrogate		
									A071204	No Pacing		
					CT983	Surgical invasive	E011903	Syncope/Fainting	A071102	Failure to Interrogate		
									A071204	No Pacing		
							E2403	No Clinical Signs, Symptoms or Conditions	A071102	Failure to Interrogate		
									A071204	No Pacing		
							F1901	Additional Surgery	A071102	Failure to Interrogate		
									A071204	No Pacing		
7137	2021	All other reportable incidents	47265	Dual-chamber implantable pacemaker, rate-responsive	CT141	Electrodes/Leads and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		

					CT145	Cardiovascular devices	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
					CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
					CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
					CT233	Surgical	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
					CT2406	Implantable	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
					CT2407	Active implantable	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
					CT2727	Active implant devices/ancillaries	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
					CT2750	Implantable cardiac electrical stimulators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
					CT2985	Active	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
					CT3001	Implantable metal	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
					CT3002	Implantable synthetic polymer	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
					CT3006	Implantable inorganic material	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
					CT314	Electrically powered	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
					CT321	Long-term surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
					CT449	Generators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
					CT981	Single use	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
					CT983	Surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
7138	2021	All other reportable incidents	47265	Dual-chamber implantable pacemaker, rate-responsive	CT141	Electrodes/Leads and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT145	Cardiovascular devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT233	Surgical	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT2406	Implantable	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		

					CT2407	Active implantable	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT2727	Active implant devices/ancillaries	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT2750	Implantable cardiac electrical stimulators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT2985	Active	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT3001	Implantable metal	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT3002	Implantable synthetic polymer	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT3006	Implantable inorganic material	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT314	Electrically powered	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT321	Long-term surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT449	Generators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT981	Single use	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT983	Surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
7139	2021	All other reportable incidents	47270	Cardiac resynchronization therapy implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	25/11/2021	25/11/2021
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	25/11/2021	25/11/2021
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	25/11/2021	25/11/2021
7140	2021	All other reportable incidents	35852	Single-chamber implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem	06/12/2021	06/12/2021
									A070504	Premature Discharge of Battery	06/12/2021	06/12/2021
							F08	Hospitalization or Prolonged Hospitalization	A0705	Battery Problem	06/12/2021	06/12/2021
									A070504	Premature Discharge of Battery	06/12/2021	06/12/2021
							F1905	Device Revision or Replacement	A0705	Battery Problem	06/12/2021	06/12/2021
									A070504	Premature Discharge of Battery	06/12/2021	06/12/2021
7141	2021	All other reportable incidents	37265	Dual-chamber implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A04	Material Integrity Problem	26/11/2021	26/11/2021
									A070504	Premature Discharge of Battery	26/11/2021	26/11/2021
									A12	Connection Problem	26/11/2021	26/11/2021
									A150207	Difficult to Remove	26/11/2021	26/11/2021
							F08	Hospitalization or Prolonged Hospitalization	A04	Material Integrity Problem	26/11/2021	26/11/2021
									A070504	Premature Discharge of Battery	26/11/2021	26/11/2021
									A12	Connection Problem	26/11/2021	26/11/2021
									A150207	Difficult to Remove	26/11/2021	26/11/2021
							F1905	Device Revision or Replacement	A04	Material Integrity Problem	26/11/2021	26/11/2021
									A070504	Premature Discharge of Battery	26/11/2021	26/11/2021
									A12	Connection Problem	26/11/2021	26/11/2021
									A150207	Difficult to Remove	26/11/2021	26/11/2021
							F1908	Prolonged Surgery	A04	Material Integrity Problem	26/11/2021	26/11/2021
									A070504	Premature Discharge of Battery	26/11/2021	26/11/2021
									A12	Connection Problem	26/11/2021	26/11/2021
									A150207	Difficult to Remove	26/11/2021	26/11/2021
7142	2021	All other reportable incidents	47265	Dual-chamber implantable pacemaker, rate-responsive	CT141	Electrodes/Leads and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	09/09/2021	09/09/2021
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	09/09/2021	09/09/2021
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	09/09/2021	09/09/2021
					CT145	Cardiovascular devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	09/09/2021	09/09/2021
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	09/09/2021	09/09/2021
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	09/09/2021	09/09/2021
					CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	09/09/2021	09/09/2021
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	09/09/2021	09/09/2021
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	09/09/2021	09/09/2021
					CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	09/09/2021	09/09/2021
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	09/09/2021	09/09/2021
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	09/09/2021	09/09/2021
					CT233	Surgical	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	09/09/2021	09/09/2021
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	09/09/2021	09/09/2021
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	09/09/2021	09/09/2021
					CT2406	Implantable	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	09/09/2021	09/09/2021
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	09/09/2021	09/09/2021
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	09/09/2021	09/09/2021
					CT2407	Active implantable	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	09/09/2021	09/09/2021
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	09/09/2021	09/09/2021
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	09/09/2021	09/09/2021
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	09/09/2021	09/09/2021
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	09/09/2021	09/09/2021
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	09/09/2021	09/09/2021
					CT2727	Active implant devices/ancillaries	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	09/09/2021	09/09/2021
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	09/09/2021	09/09/2021
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	09/09/2021	09/09/2021
					CT2750	Implantable cardiac electrical stimulators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	09/09/2021	09/09/2021
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	09/09/2021	09/09/2021
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	09/09/2021	09/09/2021
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	09/09/2021	09/09/2021
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	09/09/2021	09/09/2021
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	09/09/2021	09/09/2021
					CT2985	Active	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	09/09/2021	09/09/2021
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	09/09/2021	09/09/2021

						F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	09/09/2021	09/09/2021
					CT3001	Implantable metal	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	09/09/2021	09/09/2021
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	09/09/2021	09/09/2021
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	09/09/2021	09/09/2021
					CT3002	Implantable synthetic polymer	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	09/09/2021	09/09/2021
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	09/09/2021	09/09/2021
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	09/09/2021	09/09/2021
					CT3006	Implantable inorganic material	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	09/09/2021	09/09/2021
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	09/09/2021	09/09/2021
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	09/09/2021	09/09/2021
					CT314	Electrically powered	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	09/09/2021	09/09/2021
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	09/09/2021	09/09/2021
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	09/09/2021	09/09/2021
					CT321	Long-term surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	09/09/2021	09/09/2021
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	09/09/2021	09/09/2021
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	09/09/2021	09/09/2021
					CT449	Generators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	09/09/2021	09/09/2021
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	09/09/2021	09/09/2021
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	09/09/2021	09/09/2021
					CT981	Single use	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	09/09/2021	09/09/2021
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	09/09/2021	09/09/2021
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	09/09/2021	09/09/2021
					CT983	Surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	09/09/2021	09/09/2021
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	09/09/2021	09/09/2021
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	09/09/2021	09/09/2021
7143	2021	All other reportable incidents	37265	Dual-chamber implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A070909	Oversensing		
									A0712	Pacing Problem		
							F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A070909	Oversensing		
									A0712	Pacing Problem		
7144	2021	All other reportable incidents	37265	Dual-chamber implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A070909	Oversensing		
									A0712	Pacing Problem		
							F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A070909	Oversensing		
									A0712	Pacing Problem		
7145	2021	All other reportable incidents	37265	Dual-chamber implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	14/10/2021	14/10/2021
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	14/10/2021	14/10/2021
							F12	Serious Injury/ Illness/ Impairment	A070504	Premature Discharge of Battery	14/10/2021	14/10/2021
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	14/10/2021	14/10/2021
7146	2021	All other reportable incidents	47265	Dual-chamber implantable pacemaker, rate-responsive	CT141	Electrodes/Leads and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem	15/12/2021	15/12/2021
									A070504	Premature Discharge of Battery	15/12/2021	15/12/2021
							F08	Hospitalization or Prolonged Hospitalization	A0705	Battery Problem	15/12/2021	15/12/2021
									A070504	Premature Discharge of Battery	15/12/2021	15/12/2021
							F1905	Device Revision or Replacement	A0705	Battery Problem	15/12/2021	15/12/2021
									A070504	Premature Discharge of Battery	15/12/2021	15/12/2021
					CT145	Cardiovascular devices	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem	15/12/2021	15/12/2021
									A070504	Premature Discharge of Battery	15/12/2021	15/12/2021
							F08	Hospitalization or Prolonged Hospitalization	A0705	Battery Problem	15/12/2021	15/12/2021
									A070504	Premature Discharge of Battery	15/12/2021	15/12/2021
							F1905	Device Revision or Replacement	A0705	Battery Problem	15/12/2021	15/12/2021
									A070504	Premature Discharge of Battery	15/12/2021	15/12/2021
					CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem	15/12/2021	15/12/2021
									A070504	Premature Discharge of Battery	15/12/2021	15/12/2021
							F08	Hospitalization or Prolonged Hospitalization	A0705	Battery Problem	15/12/2021	15/12/2021
									A070504	Premature Discharge of Battery	15/12/2021	15/12/2021
							F1905	Device Revision or Replacement	A0705	Battery Problem	15/12/2021	15/12/2021
									A070504	Premature Discharge of Battery	15/12/2021	15/12/2021
					CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem	15/12/2021	15/12/2021
									A070504	Premature Discharge of Battery	15/12/2021	15/12/2021
							F08	Hospitalization or Prolonged Hospitalization	A0705	Battery Problem	15/12/2021	15/12/2021
									A070504	Premature Discharge of Battery	15/12/2021	15/12/2021
							F1905	Device Revision or Replacement	A0705	Battery Problem	15/12/2021	15/12/2021
									A070504	Premature Discharge of Battery	15/12/2021	15/12/2021
					CT233	Surgical	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem	15/12/2021	15/12/2021
									A070504	Premature Discharge of Battery	15/12/2021	15/12/2021
							F08	Hospitalization or Prolonged Hospitalization	A0705	Battery Problem	15/12/2021	15/12/2021
									A070504	Premature Discharge of Battery	15/12/2021	15/12/2021
							F1905	Device Revision or Replacement	A0705	Battery Problem	15/12/2021	15/12/2021
									A070504	Premature Discharge of Battery	15/12/2021	15/12/2021
					CT2406	Implantable	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem	15/12/2021	15/12/2021
									A070504	Premature Discharge of Battery	15/12/2021	15/12/2021
							F08	Hospitalization or Prolonged Hospitalization	A0705	Battery Problem	15/12/2021	15/12/2021
									A070504	Premature Discharge of Battery	15/12/2021	15/12/2021
							F1905	Device Revision or Replacement	A0705	Battery Problem	15/12/2021	15/12/2021
									A070504	Premature Discharge of Battery	15/12/2021	15/12/2021
					CT2407	Active implantable	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem	15/12/2021	15/12/2021
									A070504	Premature Discharge of Battery	15/12/2021	15/12/2021
							F08	Hospitalization or Prolonged Hospitalization	A0705	Battery Problem	15/12/2021	15/12/2021
									A070504	Premature Discharge of Battery	15/12/2021	15/12/2021
							F1905	Device Revision or Replacement	A0705	Battery Problem	15/12/2021	15/12/2021
									A070504	Premature Discharge of Battery	15/12/2021	15/12/2021
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem	15/12/2021	15/12/2021
									A070504	Premature Discharge of Battery	15/12/2021	15/12/2021
							F08	Hospitalization or Prolonged Hospitalization	A0705	Battery Problem	15/12/2021	15/12/2021
									A070504	Premature Discharge of Battery	15/12/2021	15/12/2021
							F1905	Device Revision or Replacement	A0705	Battery Problem	15/12/2021	15/12/2021
									A070504	Premature Discharge of Battery	15/12/2021	15/12/2021
					CT2727	Active implant devices/ancillaries	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem	15/12/2021	15/12/2021
									A070504	Premature Discharge of Battery	15/12/2021	15/12/2021
							F08	Hospitalization or Prolonged Hospitalization	A0705	Battery Problem	15/12/2021	15/12/2021

									A070504	Premature Discharge of Battery	15/12/2021	15/12/2021
							F1905	Device Revision or Replacement	A0705	Battery Problem	15/12/2021	15/12/2021
									A070504	Premature Discharge of Battery	15/12/2021	15/12/2021
					CT2750	Implantable cardiac electrical stimulators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem	15/12/2021	15/12/2021
									A070504	Premature Discharge of Battery	15/12/2021	15/12/2021
							F08	Hospitalization or Prolonged Hospitalization	A0705	Battery Problem	15/12/2021	15/12/2021
									A070504	Premature Discharge of Battery	15/12/2021	15/12/2021
							F1905	Device Revision or Replacement	A0705	Battery Problem	15/12/2021	15/12/2021
									A070504	Premature Discharge of Battery	15/12/2021	15/12/2021
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem	15/12/2021	15/12/2021
									A070504	Premature Discharge of Battery	15/12/2021	15/12/2021
							F08	Hospitalization or Prolonged Hospitalization	A0705	Battery Problem	15/12/2021	15/12/2021
									A070504	Premature Discharge of Battery	15/12/2021	15/12/2021
							F1905	Device Revision or Replacement	A0705	Battery Problem	15/12/2021	15/12/2021
									A070504	Premature Discharge of Battery	15/12/2021	15/12/2021
					CT2985	Active	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem	15/12/2021	15/12/2021
									A070504	Premature Discharge of Battery	15/12/2021	15/12/2021
							F08	Hospitalization or Prolonged Hospitalization	A0705	Battery Problem	15/12/2021	15/12/2021
									A070504	Premature Discharge of Battery	15/12/2021	15/12/2021
							F1905	Device Revision or Replacement	A0705	Battery Problem	15/12/2021	15/12/2021
									A070504	Premature Discharge of Battery	15/12/2021	15/12/2021
					CT3001	Implantable metal	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem	15/12/2021	15/12/2021
									A070504	Premature Discharge of Battery	15/12/2021	15/12/2021
							F08	Hospitalization or Prolonged Hospitalization	A0705	Battery Problem	15/12/2021	15/12/2021
									A070504	Premature Discharge of Battery	15/12/2021	15/12/2021
							F1905	Device Revision or Replacement	A0705	Battery Problem	15/12/2021	15/12/2021
									A070504	Premature Discharge of Battery	15/12/2021	15/12/2021
					CT3002	Implantable synthetic polymer	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem	15/12/2021	15/12/2021
									A070504	Premature Discharge of Battery	15/12/2021	15/12/2021
							F08	Hospitalization or Prolonged Hospitalization	A0705	Battery Problem	15/12/2021	15/12/2021
									A070504	Premature Discharge of Battery	15/12/2021	15/12/2021
							F1905	Device Revision or Replacement	A0705	Battery Problem	15/12/2021	15/12/2021
									A070504	Premature Discharge of Battery	15/12/2021	15/12/2021
					CT3006	Implantable inorganic material	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem	15/12/2021	15/12/2021
									A070504	Premature Discharge of Battery	15/12/2021	15/12/2021
							F08	Hospitalization or Prolonged Hospitalization	A0705	Battery Problem	15/12/2021	15/12/2021
									A070504	Premature Discharge of Battery	15/12/2021	15/12/2021
							F1905	Device Revision or Replacement	A0705	Battery Problem	15/12/2021	15/12/2021
									A070504	Premature Discharge of Battery	15/12/2021	15/12/2021
					CT314	Electrically powered	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem	15/12/2021	15/12/2021
									A070504	Premature Discharge of Battery	15/12/2021	15/12/2021
							F08	Hospitalization or Prolonged Hospitalization	A0705	Battery Problem	15/12/2021	15/12/2021
									A070504	Premature Discharge of Battery	15/12/2021	15/12/2021
							F1905	Device Revision or Replacement	A0705	Battery Problem	15/12/2021	15/12/2021
									A070504	Premature Discharge of Battery	15/12/2021	15/12/2021
					CT321	Long-term surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem	15/12/2021	15/12/2021
									A070504	Premature Discharge of Battery	15/12/2021	15/12/2021
							F08	Hospitalization or Prolonged Hospitalization	A0705	Battery Problem	15/12/2021	15/12/2021
									A070504	Premature Discharge of Battery	15/12/2021	15/12/2021
							F1905	Device Revision or Replacement	A0705	Battery Problem	15/12/2021	15/12/2021
									A070504	Premature Discharge of Battery	15/12/2021	15/12/2021
					CT449	Generators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem	15/12/2021	15/12/2021
									A070504	Premature Discharge of Battery	15/12/2021	15/12/2021
							F08	Hospitalization or Prolonged Hospitalization	A0705	Battery Problem	15/12/2021	15/12/2021
									A070504	Premature Discharge of Battery	15/12/2021	15/12/2021
							F1905	Device Revision or Replacement	A0705	Battery Problem	15/12/2021	15/12/2021
									A070504	Premature Discharge of Battery	15/12/2021	15/12/2021
					CT981	Single use	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem	15/12/2021	15/12/2021
									A070504	Premature Discharge of Battery	15/12/2021	15/12/2021
							F08	Hospitalization or Prolonged Hospitalization	A0705	Battery Problem	15/12/2021	15/12/2021
									A070504	Premature Discharge of Battery	15/12/2021	15/12/2021
							F1905	Device Revision or Replacement	A0705	Battery Problem	15/12/2021	15/12/2021
									A070504	Premature Discharge of Battery	15/12/2021	15/12/2021
					CT983	Surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem	15/12/2021	15/12/2021
									A070504	Premature Discharge of Battery	15/12/2021	15/12/2021
							F08	Hospitalization or Prolonged Hospitalization	A0705	Battery Problem	15/12/2021	15/12/2021
									A070504	Premature Discharge of Battery	15/12/2021	15/12/2021
							F1905	Device Revision or Replacement	A0705	Battery Problem	15/12/2021	15/12/2021
									A070504	Premature Discharge of Battery	15/12/2021	15/12/2021
7147	2021	All other reportable incidents	37265	Dual-chamber implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem	26/11/2021	26/11/2021
									A070504	Premature Discharge of Battery	26/11/2021	26/11/2021
							F08	Hospitalization or Prolonged Hospitalization	A0705	Battery Problem	26/11/2021	26/11/2021
									A070504	Premature Discharge of Battery	26/11/2021	26/11/2021
							F1905	Device Revision or Replacement	A0705	Battery Problem	26/11/2021	26/11/2021
									A070504	Premature Discharge of Battery	26/11/2021	26/11/2021
7148	2022	All other reportable incidents	17882	Physiologic-monitoring defibrillation system	CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A1603	Failure of Device to Self-Test		
							F27	Problem Identified During Non-Clinical Procedure	A1603	Failure of Device to Self-Test		
7149	2022	Death	47265	Dual-chamber implantable pacemaker, rate-responsive	CT141	Electrodes/Leads and associated devices	E2401	Insufficient Information	A24	No Apparent Device Problem	26/06/2022	26/06/2022
							F02	Death	A24	No Apparent Device Problem	26/06/2022	26/06/2022
					CT145	Cardiovascular devices	E2401	Insufficient Information	A24	No Apparent Device Problem	26/06/2022	26/06/2022
							F02	Death	A24	No Apparent Device Problem	26/06/2022	26/06/2022
					CT1684	Cardiac pacemakers and associated devices	E2401	Insufficient Information	A24	No Apparent Device Problem	26/06/2022	26/06/2022
							F02	Death	A24	No Apparent Device Problem	26/06/2022	26/06/2022
					CT1685	Cardiac pacemakers	E2401	Insufficient Information	A24	No Apparent Device Problem	26/06/2022	26/06/2022
							F02	Death	A24	No Apparent Device Problem	26/06/2022	26/06/2022
					CT233	Surgical	E2401	Insufficient Information	A24	No Apparent Device Problem	26/06/2022	26/06/2022
							F02	Death	A24	No Apparent Device Problem	26/06/2022	26/06/2022
					CT2406	Implantable	E2401	Insufficient Information	A24	No Apparent Device Problem	26/06/2022	26/06/2022

							F02	Death	A24	No Apparent Device Problem	26/06/2022	26/06/2022
					CT2407	Active implantable	E2401	Insufficient Information	A24	No Apparent Device Problem	26/06/2022	26/06/2022
							F02	Death	A24	No Apparent Device Problem	26/06/2022	26/06/2022
					CT257	Implantable cardiac pacemakers	E2401	Insufficient Information	A24	No Apparent Device Problem	26/06/2022	26/06/2022
							F02	Death	A24	No Apparent Device Problem	26/06/2022	26/06/2022
					CT2727	Active implant devices/ancillaries	E2401	Insufficient Information	A24	No Apparent Device Problem	26/06/2022	26/06/2022
							F02	Death	A24	No Apparent Device Problem	26/06/2022	26/06/2022
					CT2750	Implantable cardiac electrical stimulators and associated devices	E2401	Insufficient Information	A24	No Apparent Device Problem	26/06/2022	26/06/2022
							F02	Death	A24	No Apparent Device Problem	26/06/2022	26/06/2022
					CT2866	Implantable cardiac electrical stimulators	E2401	Insufficient Information	A24	No Apparent Device Problem	26/06/2022	26/06/2022
							F02	Death	A24	No Apparent Device Problem	26/06/2022	26/06/2022
					CT2985	Active	E2401	Insufficient Information	A24	No Apparent Device Problem	26/06/2022	26/06/2022
							F02	Death	A24	No Apparent Device Problem	26/06/2022	26/06/2022
					CT3001	Implantable metal	E2401	Insufficient Information	A24	No Apparent Device Problem	26/06/2022	26/06/2022
							F02	Death	A24	No Apparent Device Problem	26/06/2022	26/06/2022
					CT3002	Implantable synthetic polymer	E2401	Insufficient Information	A24	No Apparent Device Problem	26/06/2022	26/06/2022
							F02	Death	A24	No Apparent Device Problem	26/06/2022	26/06/2022
					CT3006	Implantable inorganic material	E2401	Insufficient Information	A24	No Apparent Device Problem	26/06/2022	26/06/2022
							F02	Death	A24	No Apparent Device Problem	26/06/2022	26/06/2022
					CT314	Electrically powered	E2401	Insufficient Information	A24	No Apparent Device Problem	26/06/2022	26/06/2022
							F02	Death	A24	No Apparent Device Problem	26/06/2022	26/06/2022
					CT321	Long-term surgical invasive	E2401	Insufficient Information	A24	No Apparent Device Problem	26/06/2022	26/06/2022
							F02	Death	A24	No Apparent Device Problem	26/06/2022	26/06/2022
					CT449	Generators and associated devices	E2401	Insufficient Information	A24	No Apparent Device Problem	26/06/2022	26/06/2022
							F02	Death	A24	No Apparent Device Problem	26/06/2022	26/06/2022
					CT981	Single use	E2401	Insufficient Information	A24	No Apparent Device Problem	26/06/2022	26/06/2022
							F02	Death	A24	No Apparent Device Problem	26/06/2022	26/06/2022
					CT983	Surgical invasive	E2401	Insufficient Information	A24	No Apparent Device Problem	26/06/2022	26/06/2022
							F02	Death	A24	No Apparent Device Problem	26/06/2022	26/06/2022
7150	2022	All other reportable incidents	47263	Cardiac resynchronization therapy implantable pacemaker	CT1684	Cardiac pacemakers and associated devices	E0717	Dyspnea	A0908	Incorrect, Inadequate or Imprecise Result or Readings		
							E2403	No Clinical Signs, Symptoms or Conditions	A0908	Incorrect, Inadequate or Imprecise Result or Readings		
							F08	Hospitalization or Prolonged Hospitalization	A0908	Incorrect, Inadequate or Imprecise Result or Readings		
							F1905	Device Revision or Replacement	A0908	Incorrect, Inadequate or Imprecise Result or Readings		
					CT1685	Cardiac pacemakers	E0717	Dyspnea	A0908	Incorrect, Inadequate or Imprecise Result or Readings		
							E2403	No Clinical Signs, Symptoms or Conditions	A0908	Incorrect, Inadequate or Imprecise Result or Readings		
							F08	Hospitalization or Prolonged Hospitalization	A0908	Incorrect, Inadequate or Imprecise Result or Readings		
							F1905	Device Revision or Replacement	A0908	Incorrect, Inadequate or Imprecise Result or Readings		
					CT257	Implantable cardiac pacemakers	E0717	Dyspnea	A0908	Incorrect, Inadequate or Imprecise Result or Readings		
							E2403	No Clinical Signs, Symptoms or Conditions	A0908	Incorrect, Inadequate or Imprecise Result or Readings		
							F08	Hospitalization or Prolonged Hospitalization	A0908	Incorrect, Inadequate or Imprecise Result or Readings		
							F1905	Device Revision or Replacement	A0908	Incorrect, Inadequate or Imprecise Result or Readings		
					CT2866	Implantable cardiac electrical stimulators	E0717	Dyspnea	A0908	Incorrect, Inadequate or Imprecise Result or Readings		
							E2403	No Clinical Signs, Symptoms or Conditions	A0908	Incorrect, Inadequate or Imprecise Result or Readings		
							F08	Hospitalization or Prolonged Hospitalization	A0908	Incorrect, Inadequate or Imprecise Result or Readings		
							F1905	Device Revision or Replacement	A0908	Incorrect, Inadequate or Imprecise Result or Readings		
7151	2022	All other reportable incidents	47263	Cardiac resynchronization therapy implantable pacemaker	CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A0908	Incorrect, Inadequate or Imprecise Result or Readings	04/07/2022	04/07/2022
							F08	Hospitalization or Prolonged Hospitalization	A0908	Incorrect, Inadequate or Imprecise Result or Readings	04/07/2022	04/07/2022
							F1905	Device Revision or Replacement	A0908	Incorrect, Inadequate or Imprecise Result or Readings	04/07/2022	04/07/2022
					CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A0908	Incorrect, Inadequate or Imprecise Result or Readings	04/07/2022	04/07/2022
							F08	Hospitalization or Prolonged Hospitalization	A0908	Incorrect, Inadequate or Imprecise Result or Readings	04/07/2022	04/07/2022
							F1905	Device Revision or Replacement	A0908	Incorrect, Inadequate or Imprecise Result or Readings	04/07/2022	04/07/2022
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A0908	Incorrect, Inadequate or Imprecise Result or Readings	04/07/2022	04/07/2022
							F08	Hospitalization or Prolonged Hospitalization	A0908	Incorrect, Inadequate or Imprecise Result or Readings	04/07/2022	04/07/2022
							F1905	Device Revision or Replacement	A0908	Incorrect, Inadequate or Imprecise Result or Readings	04/07/2022	04/07/2022
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A0908	Incorrect, Inadequate or Imprecise Result or Readings	04/07/2022	04/07/2022
							F08	Hospitalization or Prolonged Hospitalization	A0908	Incorrect, Inadequate or Imprecise Result or Readings	04/07/2022	04/07/2022
							F1905	Device Revision or Replacement	A0908	Incorrect, Inadequate or Imprecise Result or Readings	04/07/2022	04/07/2022
7152	2022	All other reportable incidents	47270	Cardiac resynchronization therapy implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A072201	High Impedance	20/06/2022	20/06/2022
							F08	Hospitalization or Prolonged Hospitalization	A072201	High Impedance	20/06/2022	20/06/2022
							F1905	Device Revision or Replacement	A072201	High Impedance	20/06/2022	20/06/2022
7153	2021	All other reportable incidents	36241	Pacemaker/defibrillator lead	CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A072201	High Impedance		
							F26	No Health Consequences or Impact	A072201	High Impedance		
7154	2021	All other reportable incidents	47265	Dual-chamber implantable pacemaker, rate-responsive	CT141	Electrodes/Leads and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT145	Cardiovascular devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		

					CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT233	Surgical	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT2406	Implantable	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT2407	Active implantable	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT2727	Active implant devices/ancillaries	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT2750	Implantable cardiac electrical stimulators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT2985	Active	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT3001	Implantable metal	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT3002	Implantable synthetic polymer	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT3006	Implantable inorganic material	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT314	Electrically powered	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT321	Long-term surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT449	Generators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT981	Single use	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT983	Surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
7155	2022	All other reportable incidents	47267	Single-chamber implantable pacemaker, rate-responsive	CT141	Electrodes/Leads and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT145	Cardiovascular devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT233	Surgical	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT2406	Implantable	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT2407	Active implantable	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT2727	Active implant devices/ancillaries	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT2750	Implantable cardiac electrical stimulators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT2985	Active	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT3001	Implantable metal	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT3002	Implantable synthetic polymer	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT3006	Implantable inorganic material	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT314	Electrically powered	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT321	Long-term surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT449	Generators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT981	Single use	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT983	Surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
7156	2022	All other reportable incidents	35852	Single-chamber implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem	04/07/2022	04/07/2022
									A070504	Premature Discharge of Battery	04/07/2022	04/07/2022
							F08	Hospitalization or Prolonged Hospitalization	A0705	Battery Problem	04/07/2022	04/07/2022
									A070504	Premature Discharge of Battery	04/07/2022	04/07/2022
							F1905	Device Revision or Replacement	A0705	Battery Problem	04/07/2022	04/07/2022
									A070504	Premature Discharge of Battery	04/07/2022	04/07/2022
7157	2021	All other reportable incidents	35852	Single-chamber implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	25/08/2021	25/08/2021
							F1901	Additional Surgery	A070504	Premature Discharge of Battery	25/08/2021	25/08/2021
7158	2021	All other reportable incidents	47267	Single-chamber implantable pacemaker, rate-responsive	CT141	Electrodes/Leads and associated devices	E0601	Arrhythmia	A071204	No Pacing	16/11/2021	16/11/2021
							E2403	No Clinical Signs, Symptoms or Conditions	A071204	No Pacing	16/11/2021	16/11/2021
							F1908	Prolonged Surgery	A071204	No Pacing	16/11/2021	16/11/2021
					CT145	Cardiovascular devices	E0601	Arrhythmia	A071204	No Pacing	16/11/2021	16/11/2021

						E2403	No Clinical Signs, Symptoms or Conditions	A071204	No Pacing	16/11/2021	16/11/2021
						F1908	Prolonged Surgery	A071204	No Pacing	16/11/2021	16/11/2021
					CT1684	Cardiac pacemakers and associated devices	E0601	Arrhythmia	A071204	No Pacing	16/11/2021	16/11/2021
							E2403	No Clinical Signs, Symptoms or Conditions	A071204	No Pacing	16/11/2021	16/11/2021
							F1908	Prolonged Surgery	A071204	No Pacing	16/11/2021	16/11/2021
					CT1685	Cardiac pacemakers	E0601	Arrhythmia	A071204	No Pacing	16/11/2021	16/11/2021
							E2403	No Clinical Signs, Symptoms or Conditions	A071204	No Pacing	16/11/2021	16/11/2021
							F1908	Prolonged Surgery	A071204	No Pacing	16/11/2021	16/11/2021
					CT233	Surgical	E0601	Arrhythmia	A071204	No Pacing	16/11/2021	16/11/2021
							E2403	No Clinical Signs, Symptoms or Conditions	A071204	No Pacing	16/11/2021	16/11/2021
							F1908	Prolonged Surgery	A071204	No Pacing	16/11/2021	16/11/2021
					CT2406	Implantable	E0601	Arrhythmia	A071204	No Pacing	16/11/2021	16/11/2021
							E2403	No Clinical Signs, Symptoms or Conditions	A071204	No Pacing	16/11/2021	16/11/2021
							F1908	Prolonged Surgery	A071204	No Pacing	16/11/2021	16/11/2021
					CT2407	Active implantable	E0601	Arrhythmia	A071204	No Pacing	16/11/2021	16/11/2021
							E2403	No Clinical Signs, Symptoms or Conditions	A071204	No Pacing	16/11/2021	16/11/2021
							F1908	Prolonged Surgery	A071204	No Pacing	16/11/2021	16/11/2021
					CT257	Implantable cardiac pacemakers	E0601	Arrhythmia	A071204	No Pacing	16/11/2021	16/11/2021
							E2403	No Clinical Signs, Symptoms or Conditions	A071204	No Pacing	16/11/2021	16/11/2021
							F1908	Prolonged Surgery	A071204	No Pacing	16/11/2021	16/11/2021
					CT2727	Active implant devices/ancillaries	E0601	Arrhythmia	A071204	No Pacing	16/11/2021	16/11/2021
							E2403	No Clinical Signs, Symptoms or Conditions	A071204	No Pacing	16/11/2021	16/11/2021
							F1908	Prolonged Surgery	A071204	No Pacing	16/11/2021	16/11/2021
					CT2750	Implantable cardiac electrical stimulators and associated devices	E0601	Arrhythmia	A071204	No Pacing	16/11/2021	16/11/2021
							E2403	No Clinical Signs, Symptoms or Conditions	A071204	No Pacing	16/11/2021	16/11/2021
							F1908	Prolonged Surgery	A071204	No Pacing	16/11/2021	16/11/2021
					CT2866	Implantable cardiac electrical stimulators	E0601	Arrhythmia	A071204	No Pacing	16/11/2021	16/11/2021
							E2403	No Clinical Signs, Symptoms or Conditions	A071204	No Pacing	16/11/2021	16/11/2021
							F1908	Prolonged Surgery	A071204	No Pacing	16/11/2021	16/11/2021
					CT2985	Active	E0601	Arrhythmia	A071204	No Pacing	16/11/2021	16/11/2021
							E2403	No Clinical Signs, Symptoms or Conditions	A071204	No Pacing	16/11/2021	16/11/2021
							F1908	Prolonged Surgery	A071204	No Pacing	16/11/2021	16/11/2021
					CT3001	Implantable metal	E0601	Arrhythmia	A071204	No Pacing	16/11/2021	16/11/2021
							E2403	No Clinical Signs, Symptoms or Conditions	A071204	No Pacing	16/11/2021	16/11/2021
							F1908	Prolonged Surgery	A071204	No Pacing	16/11/2021	16/11/2021
					CT3002	Implantable synthetic polymer	E0601	Arrhythmia	A071204	No Pacing	16/11/2021	16/11/2021
							E2403	No Clinical Signs, Symptoms or Conditions	A071204	No Pacing	16/11/2021	16/11/2021
							F1908	Prolonged Surgery	A071204	No Pacing	16/11/2021	16/11/2021
					CT3006	Implantable inorganic material	E0601	Arrhythmia	A071204	No Pacing	16/11/2021	16/11/2021
							E2403	No Clinical Signs, Symptoms or Conditions	A071204	No Pacing	16/11/2021	16/11/2021
							F1908	Prolonged Surgery	A071204	No Pacing	16/11/2021	16/11/2021
					CT314	Electrically powered	E0601	Arrhythmia	A071204	No Pacing	16/11/2021	16/11/2021
							E2403	No Clinical Signs, Symptoms or Conditions	A071204	No Pacing	16/11/2021	16/11/2021
							F1908	Prolonged Surgery	A071204	No Pacing	16/11/2021	16/11/2021
					CT321	Long-term surgical invasive	E0601	Arrhythmia	A071204	No Pacing	16/11/2021	16/11/2021
							E2403	No Clinical Signs, Symptoms or Conditions	A071204	No Pacing	16/11/2021	16/11/2021
							F1908	Prolonged Surgery	A071204	No Pacing	16/11/2021	16/11/2021
					CT449	Generators and associated devices	E0601	Arrhythmia	A071204	No Pacing	16/11/2021	16/11/2021
							E2403	No Clinical Signs, Symptoms or Conditions	A071204	No Pacing	16/11/2021	16/11/2021
							F1908	Prolonged Surgery	A071204	No Pacing	16/11/2021	16/11/2021
					CT981	Single use	E0601	Arrhythmia	A071204	No Pacing	16/11/2021	16/11/2021
							E2403	No Clinical Signs, Symptoms or Conditions	A071204	No Pacing	16/11/2021	16/11/2021
							F1908	Prolonged Surgery	A071204	No Pacing	16/11/2021	16/11/2021
					CT983	Surgical invasive	E0601	Arrhythmia	A071204	No Pacing	16/11/2021	16/11/2021
							E2403	No Clinical Signs, Symptoms or Conditions	A071204	No Pacing	16/11/2021	16/11/2021
							F1908	Prolonged Surgery	A071204	No Pacing	16/11/2021	16/11/2021
7159	2021	All other reportable incidents	47270	Cardiac resynchronization therapy implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A071302	Inappropriate/Inadequate Shock/Stimulation		
									A1305	Wireless Communication Problem		
							F23	Unexpected Medical Intervention	A071302	Inappropriate/Inadequate Shock/Stimulation		
									A1305	Wireless Communication Problem		
7160	2021	All other reportable incidents	47263	Cardiac resynchronization therapy implantable pacemaker	CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A0701	Capturing Problem	16/11/2021	16/11/2021
									A070504	Premature Discharge of Battery	16/11/2021	16/11/2021
							F1901	Additional Surgery	A0701	Capturing Problem	16/11/2021	16/11/2021
									A070504	Premature Discharge of Battery	16/11/2021	16/11/2021
					CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A0701	Capturing Problem	16/11/2021	16/11/2021
									A070504	Premature Discharge of Battery	16/11/2021	16/11/2021
							F1901	Additional Surgery	A0701	Capturing Problem	16/11/2021	16/11/2021
									A070504	Premature Discharge of Battery	16/11/2021	16/11/2021
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A0701	Capturing Problem	16/11/2021	16/11/2021
									A070504	Premature Discharge of Battery	16/11/2021	16/11/2021
							F1901	Additional Surgery	A0701	Capturing Problem	16/11/2021	16/11/2021
									A070504	Premature Discharge of Battery	16/11/2021	16/11/2021
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A0701	Capturing Problem	16/11/2021	16/11/2021
									A070504	Premature Discharge of Battery	16/11/2021	16/11/2021
							F1901	Additional Surgery	A0701	Capturing Problem	16/11/2021	16/11/2021
									A070504	Premature Discharge of Battery	16/11/2021	16/11/2021
7161	2022	All other reportable incidents	17882	Physiologic-monitoring defibrillation system	CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A071301	Failure to Deliver Shock/Stimulation		
							F27	Problem Identified During Non-Clinical Procedure	A071301	Failure to Deliver Shock/Stimulation		
7162	2021	All other reportable incidents	35223	Endocardial/interventricular septal pacing lead	CT1684	Cardiac pacemakers and associated devices	E2402	Appropriate Clinical Signs, Symptoms and Conditions Term/Code not Available	A070901	Decreased Sensitivity		
									A070909	Oversensing		
									A070910	Undersensing		
									A090801	Signal Artifact/Noise		
							E2403	No Clinical Signs, Symptoms or Conditions	A070901	Decreased Sensitivity		
									A070909	Oversensing		
									A070910	Undersensing		
									A090801	Signal Artifact/Noise		
							F28	Appropriate Health Impact Term/Code not Available	A070901	Decreased Sensitivity		
									A070909	Oversensing		

									A070910	Undersensing		
									A090801	Signal Artifact/Noise		
7163	2021	All other reportable incidents	35852	Single-chamber implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem	12/01/2022	12/01/2022
									A070504	Premature Discharge of Battery	12/01/2022	12/01/2022
							F08	Hospitalization or Prolonged Hospitalization	A0705	Battery Problem	12/01/2022	12/01/2022
									A070504	Premature Discharge of Battery	12/01/2022	12/01/2022
							F1905	Device Revision or Replacement	A0705	Battery Problem	12/01/2022	12/01/2022
									A070504	Premature Discharge of Battery	12/01/2022	12/01/2022
7164	2021	All other reportable incidents	47265	Dual-chamber implantable pacemaker, rate-responsive	CT141	Electrodes/Leads and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	01/03/2022	01/03/2022
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	01/03/2022	01/03/2022
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	01/03/2022	01/03/2022
					CT145	Cardiovascular devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	01/03/2022	01/03/2022
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	01/03/2022	01/03/2022
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	01/03/2022	01/03/2022
					CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	01/03/2022	01/03/2022
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	01/03/2022	01/03/2022
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	01/03/2022	01/03/2022
					CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	01/03/2022	01/03/2022
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	01/03/2022	01/03/2022
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	01/03/2022	01/03/2022
					CT233	Surgical	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	01/03/2022	01/03/2022
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	01/03/2022	01/03/2022
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	01/03/2022	01/03/2022
					CT2406	Implantable	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	01/03/2022	01/03/2022
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	01/03/2022	01/03/2022
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	01/03/2022	01/03/2022
					CT2407	Active implantable	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	01/03/2022	01/03/2022
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	01/03/2022	01/03/2022
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	01/03/2022	01/03/2022
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	01/03/2022	01/03/2022
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	01/03/2022	01/03/2022
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	01/03/2022	01/03/2022
					CT2727	Active implant devices/ancillaries	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	01/03/2022	01/03/2022
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	01/03/2022	01/03/2022
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	01/03/2022	01/03/2022
					CT2750	Implantable cardiac electrical stimulators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	01/03/2022	01/03/2022
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	01/03/2022	01/03/2022
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	01/03/2022	01/03/2022
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	01/03/2022	01/03/2022
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	01/03/2022	01/03/2022
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	01/03/2022	01/03/2022
					CT2985	Active	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	01/03/2022	01/03/2022
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	01/03/2022	01/03/2022
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	01/03/2022	01/03/2022
					CT3001	Implantable metal	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	01/03/2022	01/03/2022
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	01/03/2022	01/03/2022
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	01/03/2022	01/03/2022
					CT3002	Implantable synthetic polymer	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	01/03/2022	01/03/2022
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	01/03/2022	01/03/2022
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	01/03/2022	01/03/2022
					CT3006	Implantable inorganic material	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	01/03/2022	01/03/2022
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	01/03/2022	01/03/2022
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	01/03/2022	01/03/2022
					CT314	Electrically powered	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	01/03/2022	01/03/2022
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	01/03/2022	01/03/2022
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	01/03/2022	01/03/2022
					CT321	Long-term surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	01/03/2022	01/03/2022
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	01/03/2022	01/03/2022
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	01/03/2022	01/03/2022
					CT449	Generators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	01/03/2022	01/03/2022
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	01/03/2022	01/03/2022
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	01/03/2022	01/03/2022
					CT981	Single use	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	01/03/2022	01/03/2022
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	01/03/2022	01/03/2022
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	01/03/2022	01/03/2022
					CT983	Surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	01/03/2022	01/03/2022
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	01/03/2022	01/03/2022
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	01/03/2022	01/03/2022
7165	2021	All other reportable incidents	36241	Pacemaker/defibrillator lead	CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070101	Failure to Capture	21/10/2021	21/10/2021
									A070102	High Capture Threshold	21/10/2021	21/10/2021
									A070909	Oversensing	21/10/2021	21/10/2021
									A090801	Signal Artifact/Noise	21/10/2021	21/10/2021
							F08	Hospitalization or Prolonged Hospitalization	A070101	Failure to Capture	21/10/2021	21/10/2021
									A070102	High Capture Threshold	21/10/2021	21/10/2021
									A070909	Oversensing	21/10/2021	21/10/2021
									A090801	Signal Artifact/Noise	21/10/2021	21/10/2021
							F1905	Device Revision or Replacement	A070101	Failure to Capture	21/10/2021	21/10/2021
									A070102	High Capture Threshold	21/10/2021	21/10/2021
									A070909	Oversensing	21/10/2021	21/10/2021
									A090801	Signal Artifact/Noise	21/10/2021	21/10/2021
7166	2022	All other reportable incidents	47265	Dual-chamber implantable pacemaker, rate-responsive	CT141	Electrodes/Leads and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A071102	Failure to Interrogate	02/07/2022	02/07/2022
							F1901	Additional Surgery	A071102	Failure to Interrogate	02/07/2022	02/07/2022
					CT145	Cardiovascular devices	E2403	No Clinical Signs, Symptoms or Conditions	A071102	Failure to Interrogate	02/07/2022	02/07/2022
							F1901	Additional Surgery	A071102	Failure to Interrogate	02/07/2022	02/07/2022
					CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A071102	Failure to Interrogate	02/07/2022	02/07/2022
							F1901	Additional Surgery	A071102	Failure to Interrogate	02/07/2022	02/07/2022
					CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A071102	Failure to Interrogate	02/07/2022	02/07/2022
							F1901	Additional Surgery	A071102	Failure to Interrogate	02/07/2022	02/07/2022

					CT233	Surgical	E2403	No Clinical Signs, Symptoms or Conditions	A071102	Failure to Interrogate	02/07/2022	02/07/2022
							F1901	Additional Surgery	A071102	Failure to Interrogate	02/07/2022	02/07/2022
					CT2406	Implantable	E2403	No Clinical Signs, Symptoms or Conditions	A071102	Failure to Interrogate	02/07/2022	02/07/2022
							F1901	Additional Surgery	A071102	Failure to Interrogate	02/07/2022	02/07/2022
					CT2407	Active implantable	E2403	No Clinical Signs, Symptoms or Conditions	A071102	Failure to Interrogate	02/07/2022	02/07/2022
							F1901	Additional Surgery	A071102	Failure to Interrogate	02/07/2022	02/07/2022
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A071102	Failure to Interrogate	02/07/2022	02/07/2022
							F1901	Additional Surgery	A071102	Failure to Interrogate	02/07/2022	02/07/2022
					CT2727	Active implant devices/ancillaries	E2403	No Clinical Signs, Symptoms or Conditions	A071102	Failure to Interrogate	02/07/2022	02/07/2022
							F1901	Additional Surgery	A071102	Failure to Interrogate	02/07/2022	02/07/2022
					CT2750	Implantable cardiac electrical stimulators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A071102	Failure to Interrogate	02/07/2022	02/07/2022
							F1901	Additional Surgery	A071102	Failure to Interrogate	02/07/2022	02/07/2022
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A071102	Failure to Interrogate	02/07/2022	02/07/2022
							F1901	Additional Surgery	A071102	Failure to Interrogate	02/07/2022	02/07/2022
					CT2985	Active	E2403	No Clinical Signs, Symptoms or Conditions	A071102	Failure to Interrogate	02/07/2022	02/07/2022
							F1901	Additional Surgery	A071102	Failure to Interrogate	02/07/2022	02/07/2022
					CT3001	Implantable metal	E2403	No Clinical Signs, Symptoms or Conditions	A071102	Failure to Interrogate	02/07/2022	02/07/2022
							F1901	Additional Surgery	A071102	Failure to Interrogate	02/07/2022	02/07/2022
					CT3002	Implantable synthetic polymer	E2403	No Clinical Signs, Symptoms or Conditions	A071102	Failure to Interrogate	02/07/2022	02/07/2022
							F1901	Additional Surgery	A071102	Failure to Interrogate	02/07/2022	02/07/2022
					CT3006	Implantable inorganic material	E2403	No Clinical Signs, Symptoms or Conditions	A071102	Failure to Interrogate	02/07/2022	02/07/2022
							F1901	Additional Surgery	A071102	Failure to Interrogate	02/07/2022	02/07/2022
					CT314	Electrically powered	E2403	No Clinical Signs, Symptoms or Conditions	A071102	Failure to Interrogate	02/07/2022	02/07/2022
							F1901	Additional Surgery	A071102	Failure to Interrogate	02/07/2022	02/07/2022
					CT321	Long-term surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A071102	Failure to Interrogate	02/07/2022	02/07/2022
							F1901	Additional Surgery	A071102	Failure to Interrogate	02/07/2022	02/07/2022
					CT449	Generators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A071102	Failure to Interrogate	02/07/2022	02/07/2022
							F1901	Additional Surgery	A071102	Failure to Interrogate	02/07/2022	02/07/2022
					CT981	Single use	E2403	No Clinical Signs, Symptoms or Conditions	A071102	Failure to Interrogate	02/07/2022	02/07/2022
							F1901	Additional Surgery	A071102	Failure to Interrogate	02/07/2022	02/07/2022
					CT983	Surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A071102	Failure to Interrogate	02/07/2022	02/07/2022
							F1901	Additional Surgery	A071102	Failure to Interrogate	02/07/2022	02/07/2022
7167	2022	All other reportable incidents	35223	Endocardial/interventricular septal pacing lead	CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070909	Oversensing		
							F23	Unexpected Medical Intervention	A070909	Oversensing		
7168	2021	All other reportable incidents	47263	Cardiac resynchronization therapy implantable pacemaker	CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A0908	Incorrect, Inadequate or Imprecise Result or Readings	09/11/2021	09/11/2021
							F08	Hospitalization or Prolonged Hospitalization	A0908	Incorrect, Inadequate or Imprecise Result or Readings	09/11/2021	09/11/2021
							F1905	Device Revision or Replacement	A0908	Incorrect, Inadequate or Imprecise Result or Readings	09/11/2021	09/11/2021
					CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A0908	Incorrect, Inadequate or Imprecise Result or Readings	09/11/2021	09/11/2021
							F08	Hospitalization or Prolonged Hospitalization	A0908	Incorrect, Inadequate or Imprecise Result or Readings	09/11/2021	09/11/2021
							F1905	Device Revision or Replacement	A0908	Incorrect, Inadequate or Imprecise Result or Readings	09/11/2021	09/11/2021
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A0908	Incorrect, Inadequate or Imprecise Result or Readings	09/11/2021	09/11/2021
							F08	Hospitalization or Prolonged Hospitalization	A0908	Incorrect, Inadequate or Imprecise Result or Readings	09/11/2021	09/11/2021
							F1905	Device Revision or Replacement	A0908	Incorrect, Inadequate or Imprecise Result or Readings	09/11/2021	09/11/2021
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A0908	Incorrect, Inadequate or Imprecise Result or Readings	09/11/2021	09/11/2021
							F08	Hospitalization or Prolonged Hospitalization	A0908	Incorrect, Inadequate or Imprecise Result or Readings	09/11/2021	09/11/2021
							F1905	Device Revision or Replacement	A0908	Incorrect, Inadequate or Imprecise Result or Readings	09/11/2021	09/11/2021
7169	2021	All other reportable incidents	35852	Single-chamber implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	15/11/2021	15/11/2021
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	15/11/2021	15/11/2021
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	15/11/2021	15/11/2021
7170	2022	All other reportable incidents	36241	Pacemaker/defibrillator lead	CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A072201	High Impedance		
							F26	No Health Consequences or Impact	A072201	High Impedance		
7171	2021	All other reportable incidents	35852	Single-chamber implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	15/11/2021	15/11/2021
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	15/11/2021	15/11/2021
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	15/11/2021	15/11/2021
7172	2021	All other reportable incidents	17882	Physiologic-monitoring defibrillation system	CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A1603	Failure of Device to Self-Test		
							F27	Problem Identified During Non-Clinical Procedure	A1603	Failure of Device to Self-Test		
7173	2022	All other reportable incidents	35223	Endocardial/interventricular septal pacing lead	CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A072201	High Impedance		
							F26	No Health Consequences or Impact	A072201	High Impedance		
7174	2022	All other reportable incidents	17882	Physiologic-monitoring defibrillation system	CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070908	Failure to Sense		
							F27	Problem Identified During Non-Clinical Procedure	A070908	Failure to Sense		
7175	2022	All other reportable incidents	47265	Dual-chamber implantable pacemaker, rate-responsive	CT141	Electrodes/Leads and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	23/06/2022	23/06/2022
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	23/06/2022	23/06/2022
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	23/06/2022	23/06/2022
					CT145	Cardiovascular devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	23/06/2022	23/06/2022
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	23/06/2022	23/06/2022
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	23/06/2022	23/06/2022
					CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	23/06/2022	23/06/2022
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	23/06/2022	23/06/2022
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	23/06/2022	23/06/2022
					CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	23/06/2022	23/06/2022
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	23/06/2022	23/06/2022
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	23/06/2022	23/06/2022
					CT233	Surgical	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	23/06/2022	23/06/2022
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	23/06/2022	23/06/2022
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	23/06/2022	23/06/2022
					CT2406	Implantable	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	23/06/2022	23/06/2022
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	23/06/2022	23/06/2022
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	23/06/2022	23/06/2022
					CT2407	Active implantable	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	23/06/2022	23/06/2022

						F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	23/06/2022	23/06/2022
						F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	23/06/2022	23/06/2022
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	23/06/2022	23/06/2022
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	23/06/2022	23/06/2022
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	23/06/2022	23/06/2022
					CT2727	Active implant devices/ancillaries	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	23/06/2022	23/06/2022
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	23/06/2022	23/06/2022
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	23/06/2022	23/06/2022
					CT2750	Implantable cardiac electrical stimulators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	23/06/2022	23/06/2022
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	23/06/2022	23/06/2022
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	23/06/2022	23/06/2022
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	23/06/2022	23/06/2022
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	23/06/2022	23/06/2022
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	23/06/2022	23/06/2022
					CT2985	Active	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	23/06/2022	23/06/2022
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	23/06/2022	23/06/2022
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	23/06/2022	23/06/2022
					CT3001	Implantable metal	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	23/06/2022	23/06/2022
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	23/06/2022	23/06/2022
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	23/06/2022	23/06/2022
					CT3002	Implantable synthetic polymer	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	23/06/2022	23/06/2022
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	23/06/2022	23/06/2022
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	23/06/2022	23/06/2022
					CT3006	Implantable inorganic material	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	23/06/2022	23/06/2022
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	23/06/2022	23/06/2022
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	23/06/2022	23/06/2022
					CT314	Electrically powered	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	23/06/2022	23/06/2022
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	23/06/2022	23/06/2022
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	23/06/2022	23/06/2022
					CT321	Long-term surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	23/06/2022	23/06/2022
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	23/06/2022	23/06/2022
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	23/06/2022	23/06/2022
					CT449	Generators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	23/06/2022	23/06/2022
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	23/06/2022	23/06/2022
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	23/06/2022	23/06/2022
					CT981	Single use	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	23/06/2022	23/06/2022
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	23/06/2022	23/06/2022
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	23/06/2022	23/06/2022
					CT983	Surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	23/06/2022	23/06/2022
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	23/06/2022	23/06/2022
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	23/06/2022	23/06/2022
7176	2022	All other reportable incidents	36241	Pacemaker/defibrillator lead	CT1684	Cardiac pacemakers and associated devices	E2104	Electric Shock	A070909	Oversensing	22/06/2022	22/06/2022
									A071302	Inappropriate/Inadequate Shock/Stimulation	22/06/2022	22/06/2022
							E2403	No Clinical Signs, Symptoms or Conditions	A070909	Oversensing	22/06/2022	22/06/2022
									A071302	Inappropriate/Inadequate Shock/Stimulation	22/06/2022	22/06/2022
							F08	Hospitalization or Prolonged Hospitalization	A070909	Oversensing	22/06/2022	22/06/2022
									A071302	Inappropriate/Inadequate Shock/Stimulation	22/06/2022	22/06/2022
							F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A070909	Oversensing	22/06/2022	22/06/2022
									A071302	Inappropriate/Inadequate Shock/Stimulation	22/06/2022	22/06/2022
							F1905	Device Revision or Replacement	A070909	Oversensing	22/06/2022	22/06/2022
									A071302	Inappropriate/Inadequate Shock/Stimulation	22/06/2022	22/06/2022
7177	2021	All other reportable incidents	35852	Single-chamber implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A070602	Delayed Charge Time	02/11/2021	02/11/2021
									A0908	Incorrect, Inadequate or Imprecise Result or Readings	02/11/2021	02/11/2021
							F08	Hospitalization or Prolonged Hospitalization	A070602	Delayed Charge Time	02/11/2021	02/11/2021
									A0908	Incorrect, Inadequate or Imprecise Result or Readings	02/11/2021	02/11/2021
							F1905	Device Revision or Replacement	A070602	Delayed Charge Time	02/11/2021	02/11/2021
									A0908	Incorrect, Inadequate or Imprecise Result or Readings	02/11/2021	02/11/2021
7178	2021	All other reportable incidents	35852	Single-chamber implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E2104	Electric Shock	A070909	Oversensing		
									A071302	Inappropriate/Inadequate Shock/Stimulation		
							E2403	No Clinical Signs, Symptoms or Conditions	A070909	Oversensing		
									A071302	Inappropriate/Inadequate Shock/Stimulation		
							F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A070909	Oversensing		
									A071302	Inappropriate/Inadequate Shock/Stimulation		
7179	2021	All other reportable incidents	47267	Single-chamber implantable pacemaker, rate-responsive	CT141	Electrodes/Leads and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	25/01/2022	25/01/2022
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	25/01/2022	25/01/2022
							F1903	Device Explantation	A070504	Premature Discharge of Battery	25/01/2022	25/01/2022
					CT145	Cardiovascular devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	25/01/2022	25/01/2022
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	25/01/2022	25/01/2022
							F1903	Device Explantation	A070504	Premature Discharge of Battery	25/01/2022	25/01/2022
					CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	25/01/2022	25/01/2022
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	25/01/2022	25/01/2022
							F1903	Device Explantation	A070504	Premature Discharge of Battery	25/01/2022	25/01/2022
					CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	25/01/2022	25/01/2022
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	25/01/2022	25/01/2022
							F1903	Device Explantation	A070504	Premature Discharge of Battery	25/01/2022	25/01/2022
					CT233	Surgical	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	25/01/2022	25/01/2022
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	25/01/2022	25/01/2022
							F1903	Device Explantation	A070504	Premature Discharge of Battery	25/01/2022	25/01/2022
					CT2406	Implantable	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	25/01/2022	25/01/2022
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	25/01/2022	25/01/2022
							F1903	Device Explantation	A070504	Premature Discharge of Battery	25/01/2022	25/01/2022
					CT2407	Active implantable	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	25/01/2022	25/01/2022
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	25/01/2022	25/01/2022
							F1903	Device Explantation	A070504	Premature Discharge of Battery	25/01/2022	25/01/2022
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	25/01/2022	25/01/2022

							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	25/01/2022	25/01/2022
							F1903	Device Explantation	A070504	Premature Discharge of Battery	25/01/2022	25/01/2022
					CT2727	Active implant devices/ancillaries	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	25/01/2022	25/01/2022
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	25/01/2022	25/01/2022
							F1903	Device Explantation	A070504	Premature Discharge of Battery	25/01/2022	25/01/2022
					CT2750	Implantable cardiac electrical stimulators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	25/01/2022	25/01/2022
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	25/01/2022	25/01/2022
							F1903	Device Explantation	A070504	Premature Discharge of Battery	25/01/2022	25/01/2022
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	25/01/2022	25/01/2022
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	25/01/2022	25/01/2022
							F1903	Device Explantation	A070504	Premature Discharge of Battery	25/01/2022	25/01/2022
					CT2985	Active	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	25/01/2022	25/01/2022
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	25/01/2022	25/01/2022
							F1903	Device Explantation	A070504	Premature Discharge of Battery	25/01/2022	25/01/2022
					CT3001	Implantable metal	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	25/01/2022	25/01/2022
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	25/01/2022	25/01/2022
							F1903	Device Explantation	A070504	Premature Discharge of Battery	25/01/2022	25/01/2022
					CT3002	Implantable synthetic polymer	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	25/01/2022	25/01/2022
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	25/01/2022	25/01/2022
							F1903	Device Explantation	A070504	Premature Discharge of Battery	25/01/2022	25/01/2022
					CT3006	Implantable inorganic material	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	25/01/2022	25/01/2022
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	25/01/2022	25/01/2022
							F1903	Device Explantation	A070504	Premature Discharge of Battery	25/01/2022	25/01/2022
					CT314	Electrically powered	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	25/01/2022	25/01/2022
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	25/01/2022	25/01/2022
							F1903	Device Explantation	A070504	Premature Discharge of Battery	25/01/2022	25/01/2022
					CT321	Long-term surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	25/01/2022	25/01/2022
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	25/01/2022	25/01/2022
							F1903	Device Explantation	A070504	Premature Discharge of Battery	25/01/2022	25/01/2022
					CT449	Generators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	25/01/2022	25/01/2022
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	25/01/2022	25/01/2022
							F1903	Device Explantation	A070504	Premature Discharge of Battery	25/01/2022	25/01/2022
					CT981	Single use	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	25/01/2022	25/01/2022
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	25/01/2022	25/01/2022
							F1903	Device Explantation	A070504	Premature Discharge of Battery	25/01/2022	25/01/2022
					CT983	Surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	25/01/2022	25/01/2022
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	25/01/2022	25/01/2022
							F1903	Device Explantation	A070504	Premature Discharge of Battery	25/01/2022	25/01/2022
7180	2021	All other reportable incidents	47265	Dual-chamber implantable pacemaker, rate-responsive	CT141	Electrodes/Leads and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT145	Cardiovascular devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT233	Surgical	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT2406	Implantable	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT2407	Active implantable	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT2727	Active implant devices/ancillaries	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT2750	Implantable cardiac electrical stimulators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT2985	Active	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT3001	Implantable metal	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT3002	Implantable synthetic polymer	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT3006	Implantable inorganic material	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT314	Electrically powered	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT321	Long-term surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT449	Generators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT981	Single use	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT983	Surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
7181	2021	All other reportable incidents	60789	Intracardiac pacemaker	CT130	Catheters and associated devices	E0602	Cardiac Arrest	A24	No Apparent Device Problem		
							E0605	Cardiac Tamponade	A24	No Apparent Device Problem		
							E0619	Pericardial Effusion	A24	No Apparent Device Problem		
							E2330	Pain	A24	No Apparent Device Problem		
							E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem		
							F08	Hospitalization or Prolonged Hospitalization	A24	No Apparent Device Problem		
							F12	Serious Injury/ Illness/ Impairment	A24	No Apparent Device Problem		
							F1203	Life Threatening Illness or Injury	A24	No Apparent Device Problem		
					CT145	Cardiovascular devices	E0602	Cardiac Arrest	A24	No Apparent Device Problem		
							E0605	Cardiac Tamponade	A24	No Apparent Device Problem		

							E0619	Pericardial Effusion	A24	No Apparent Device Problem		
							E2330	Pain	A24	No Apparent Device Problem		
							E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem		
							F08	Hospitalization or Prolonged Hospitalization	A24	No Apparent Device Problem		
							F12	Serious Injury/ Illness/ Impairment	A24	No Apparent Device Problem		
							F1203	Life Threatening Illness or Injury	A24	No Apparent Device Problem		
					CT1684	Cardiac pacemakers and associated devices	E0602	Cardiac Arrest	A24	No Apparent Device Problem		
							E0605	Cardiac Tamponade	A24	No Apparent Device Problem		
							E0619	Pericardial Effusion	A24	No Apparent Device Problem		
							E2330	Pain	A24	No Apparent Device Problem		
							E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem		
							F08	Hospitalization or Prolonged Hospitalization	A24	No Apparent Device Problem		
							F12	Serious Injury/ Illness/ Impairment	A24	No Apparent Device Problem		
							F1203	Life Threatening Illness or Injury	A24	No Apparent Device Problem		
					CT1685	Cardiac pacemakers	E0602	Cardiac Arrest	A24	No Apparent Device Problem		
							E0605	Cardiac Tamponade	A24	No Apparent Device Problem		
							E0619	Pericardial Effusion	A24	No Apparent Device Problem		
							E2330	Pain	A24	No Apparent Device Problem		
							E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem		
							F08	Hospitalization or Prolonged Hospitalization	A24	No Apparent Device Problem		
							F12	Serious Injury/ Illness/ Impairment	A24	No Apparent Device Problem		
							F1203	Life Threatening Illness or Injury	A24	No Apparent Device Problem		
					CT233	Surgical	E0602	Cardiac Arrest	A24	No Apparent Device Problem		
							E0605	Cardiac Tamponade	A24	No Apparent Device Problem		
							E0619	Pericardial Effusion	A24	No Apparent Device Problem		
							E2330	Pain	A24	No Apparent Device Problem		
							E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem		
							F08	Hospitalization or Prolonged Hospitalization	A24	No Apparent Device Problem		
							F12	Serious Injury/ Illness/ Impairment	A24	No Apparent Device Problem		
							F1203	Life Threatening Illness or Injury	A24	No Apparent Device Problem		
					CT2406	Implantable	E0602	Cardiac Arrest	A24	No Apparent Device Problem		
							E0605	Cardiac Tamponade	A24	No Apparent Device Problem		
							E0619	Pericardial Effusion	A24	No Apparent Device Problem		
							E2330	Pain	A24	No Apparent Device Problem		
							E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem		
							F08	Hospitalization or Prolonged Hospitalization	A24	No Apparent Device Problem		
							F12	Serious Injury/ Illness/ Impairment	A24	No Apparent Device Problem		
							F1203	Life Threatening Illness or Injury	A24	No Apparent Device Problem		
					CT2407	Active implantable	E0602	Cardiac Arrest	A24	No Apparent Device Problem		
							E0605	Cardiac Tamponade	A24	No Apparent Device Problem		
							E0619	Pericardial Effusion	A24	No Apparent Device Problem		
							E2330	Pain	A24	No Apparent Device Problem		
							E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem		
							F08	Hospitalization or Prolonged Hospitalization	A24	No Apparent Device Problem		
							F12	Serious Injury/ Illness/ Impairment	A24	No Apparent Device Problem		
							F1203	Life Threatening Illness or Injury	A24	No Apparent Device Problem		
					CT257	Implantable cardiac pacemakers	E0602	Cardiac Arrest	A24	No Apparent Device Problem		
							E0605	Cardiac Tamponade	A24	No Apparent Device Problem		
							E0619	Pericardial Effusion	A24	No Apparent Device Problem		
							E2330	Pain	A24	No Apparent Device Problem		
							E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem		
							F08	Hospitalization or Prolonged Hospitalization	A24	No Apparent Device Problem		
							F12	Serious Injury/ Illness/ Impairment	A24	No Apparent Device Problem		
							F1203	Life Threatening Illness or Injury	A24	No Apparent Device Problem		
					CT2727	Active implant devices/ancillaries	E0602	Cardiac Arrest	A24	No Apparent Device Problem		
							E0605	Cardiac Tamponade	A24	No Apparent Device Problem		
							E0619	Pericardial Effusion	A24	No Apparent Device Problem		
							E2330	Pain	A24	No Apparent Device Problem		
							E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem		
							F08	Hospitalization or Prolonged Hospitalization	A24	No Apparent Device Problem		
							F12	Serious Injury/ Illness/ Impairment	A24	No Apparent Device Problem		
							F1203	Life Threatening Illness or Injury	A24	No Apparent Device Problem		
					CT2750	Implantable cardiac electrical stimulators and associated devices	E0602	Cardiac Arrest	A24	No Apparent Device Problem		
							E0605	Cardiac Tamponade	A24	No Apparent Device Problem		
							E0619	Pericardial Effusion	A24	No Apparent Device Problem		
							E2330	Pain	A24	No Apparent Device Problem		
							E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem		
							F08	Hospitalization or Prolonged Hospitalization	A24	No Apparent Device Problem		
							F12	Serious Injury/ Illness/ Impairment	A24	No Apparent Device Problem		
							F1203	Life Threatening Illness or Injury	A24	No Apparent Device Problem		
					CT2818	Catheterization/Cannulation/Fluid path devices	E0602	Cardiac Arrest	A24	No Apparent Device Problem		
							E0605	Cardiac Tamponade	A24	No Apparent Device Problem		
							E0619	Pericardial Effusion	A24	No Apparent Device Problem		
							E2330	Pain	A24	No Apparent Device Problem		
							E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem		
							F08	Hospitalization or Prolonged Hospitalization	A24	No Apparent Device Problem		
							F12	Serious Injury/ Illness/ Impairment	A24	No Apparent Device Problem		
							F1203	Life Threatening Illness or Injury	A24	No Apparent Device Problem		
					CT2866	Implantable cardiac electrical stimulators	E0602	Cardiac Arrest	A24	No Apparent Device Problem		
							E0605	Cardiac Tamponade	A24	No Apparent Device Problem		
							E0619	Pericardial Effusion	A24	No Apparent Device Problem		
							E2330	Pain	A24	No Apparent Device Problem		
							E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem		
							F08	Hospitalization or Prolonged Hospitalization	A24	No Apparent Device Problem		
							F12	Serious Injury/ Illness/ Impairment	A24	No Apparent Device Problem		
							F1203	Life Threatening Illness or Injury	A24	No Apparent Device Problem		
					CT2985	Active	E0602	Cardiac Arrest	A24	No Apparent Device Problem		
							E0605	Cardiac Tamponade	A24	No Apparent Device Problem		

							E0619	Pericardial Effusion	A24	No Apparent Device Problem		
							E2330	Pain	A24	No Apparent Device Problem		
							E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem		
							F08	Hospitalization or Prolonged Hospitalization	A24	No Apparent Device Problem		
							F12	Serious Injury/ Illness/ Impairment	A24	No Apparent Device Problem		
							F1203	Life Threatening Illness or Injury	A24	No Apparent Device Problem		
					CT3001	Implantable metal	E0602	Cardiac Arrest	A24	No Apparent Device Problem		
							E0605	Cardiac Tamponade	A24	No Apparent Device Problem		
							E0619	Pericardial Effusion	A24	No Apparent Device Problem		
							E2330	Pain	A24	No Apparent Device Problem		
							E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem		
							F08	Hospitalization or Prolonged Hospitalization	A24	No Apparent Device Problem		
							F12	Serious Injury/ Illness/ Impairment	A24	No Apparent Device Problem		
							F1203	Life Threatening Illness or Injury	A24	No Apparent Device Problem		
					CT3002	Implantable synthetic polymer	E0602	Cardiac Arrest	A24	No Apparent Device Problem		
							E0605	Cardiac Tamponade	A24	No Apparent Device Problem		
							E0619	Pericardial Effusion	A24	No Apparent Device Problem		
							E2330	Pain	A24	No Apparent Device Problem		
							E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem		
							F08	Hospitalization or Prolonged Hospitalization	A24	No Apparent Device Problem		
							F12	Serious Injury/ Illness/ Impairment	A24	No Apparent Device Problem		
							F1203	Life Threatening Illness or Injury	A24	No Apparent Device Problem		
					CT3006	Implantable inorganic material	E0602	Cardiac Arrest	A24	No Apparent Device Problem		
							E0605	Cardiac Tamponade	A24	No Apparent Device Problem		
							E0619	Pericardial Effusion	A24	No Apparent Device Problem		
							E2330	Pain	A24	No Apparent Device Problem		
							E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem		
							F08	Hospitalization or Prolonged Hospitalization	A24	No Apparent Device Problem		
							F12	Serious Injury/ Illness/ Impairment	A24	No Apparent Device Problem		
							F1203	Life Threatening Illness or Injury	A24	No Apparent Device Problem		
					CT314	Electrically powered	E0602	Cardiac Arrest	A24	No Apparent Device Problem		
							E0605	Cardiac Tamponade	A24	No Apparent Device Problem		
							E0619	Pericardial Effusion	A24	No Apparent Device Problem		
							E2330	Pain	A24	No Apparent Device Problem		
							E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem		
							F08	Hospitalization or Prolonged Hospitalization	A24	No Apparent Device Problem		
							F12	Serious Injury/ Illness/ Impairment	A24	No Apparent Device Problem		
							F1203	Life Threatening Illness or Injury	A24	No Apparent Device Problem		
					CT321	Long-term surgical invasive	E0602	Cardiac Arrest	A24	No Apparent Device Problem		
							E0605	Cardiac Tamponade	A24	No Apparent Device Problem		
							E0619	Pericardial Effusion	A24	No Apparent Device Problem		
							E2330	Pain	A24	No Apparent Device Problem		
							E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem		
							F08	Hospitalization or Prolonged Hospitalization	A24	No Apparent Device Problem		
							F12	Serious Injury/ Illness/ Impairment	A24	No Apparent Device Problem		
							F1203	Life Threatening Illness or Injury	A24	No Apparent Device Problem		
					CT981	Single use	E0602	Cardiac Arrest	A24	No Apparent Device Problem		
							E0605	Cardiac Tamponade	A24	No Apparent Device Problem		
							E0619	Pericardial Effusion	A24	No Apparent Device Problem		
							E2330	Pain	A24	No Apparent Device Problem		
							E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem		
							F08	Hospitalization or Prolonged Hospitalization	A24	No Apparent Device Problem		
							F12	Serious Injury/ Illness/ Impairment	A24	No Apparent Device Problem		
							F1203	Life Threatening Illness or Injury	A24	No Apparent Device Problem		
					CT983	Surgical invasive	E0602	Cardiac Arrest	A24	No Apparent Device Problem		
							E0605	Cardiac Tamponade	A24	No Apparent Device Problem		
							E0619	Pericardial Effusion	A24	No Apparent Device Problem		
							E2330	Pain	A24	No Apparent Device Problem		
							E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem		
							F08	Hospitalization or Prolonged Hospitalization	A24	No Apparent Device Problem		
							F12	Serious Injury/ Illness/ Impairment	A24	No Apparent Device Problem		
							F1203	Life Threatening Illness or Injury	A24	No Apparent Device Problem		
7182	2021	All other reportable incidents	17882	Physiologic-monitoring defibrillation system	CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070912	Incorrect Interpretation of Signal		
							F26	No Health Consequences or Impact	A070912	Incorrect Interpretation of Signal		
7183	2021	All other reportable incidents	17882	Physiologic-monitoring defibrillation system	CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070908	Failure to Sense		
							F26	No Health Consequences or Impact	A070908	Failure to Sense		
7184	2021	All other reportable incidents	17882	Physiologic-monitoring defibrillation system	CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070908	Failure to Sense		
							F26	No Health Consequences or Impact	A070908	Failure to Sense		
7185	2021	All other reportable incidents	35852	Single-chamber implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem	30/11/2021	30/11/2021
									A070504	Premature Discharge of Battery	30/11/2021	30/11/2021
							F08	Hospitalization or Prolonged Hospitalization	A0705	Battery Problem	30/11/2021	30/11/2021
									A070504	Premature Discharge of Battery	30/11/2021	30/11/2021
							F1905	Device Revision or Replacement	A0705	Battery Problem	30/11/2021	30/11/2021
									A070504	Premature Discharge of Battery	30/11/2021	30/11/2021
7186	2021	All other reportable incidents	47265	Dual-chamber implantable pacemaker, rate-responsive	CT141	Electrodes/Leads and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
					CT145	Cardiovascular devices	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
					CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
					CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		

							F26	No Health Consequences or Impact	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
					CT233	Surgical	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
					CT2406	Implantable	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
					CT2407	Active implantable	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
					CT2727	Active implant devices/ancillaries	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
					CT2750	Implantable cardiac electrical stimulators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
					CT2985	Active	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
					CT3001	Implantable metal	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
					CT3002	Implantable synthetic polymer	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
					CT3006	Implantable inorganic material	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
					CT314	Electrically powered	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
					CT321	Long-term surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
					CT449	Generators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
					CT981	Single use	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
					CT983	Surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
7187	2021	All other reportable incidents	35852	Single-chamber implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
7188	2021	All other reportable incidents	36241	Pacemaker/defibrillator lead	CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A040101	Fracture		
									A070909	Oversensing		
									A072201	High Impedance		
							F26	No Health Consequences or Impact	A040101	Fracture		
									A070909	Oversensing		
									A072201	High Impedance		
7189	2021	All other reportable incidents	47265	Dual-chamber implantable pacemaker, rate-responsive	CT141	Electrodes/Leads and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem	09/12/2021	09/12/2021
									A070504	Premature Discharge of Battery	09/12/2021	09/12/2021
									A070909	Oversensing	09/12/2021	09/12/2021
							F08	Hospitalization or Prolonged Hospitalization	A0705	Battery Problem	09/12/2021	09/12/2021
									A070504	Premature Discharge of Battery	09/12/2021	09/12/2021
									A070909	Oversensing	09/12/2021	09/12/2021
							F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A0705	Battery Problem	09/12/2021	09/12/2021
									A070504	Premature Discharge of Battery	09/12/2021	09/12/2021
									A070909	Oversensing	09/12/2021	09/12/2021
							F1905	Device Revision or Replacement	A0705	Battery Problem	09/12/2021	09/12/2021
									A070504	Premature Discharge of Battery	09/12/2021	09/12/2021
									A070909	Oversensing	09/12/2021	09/12/2021
					CT145	Cardiovascular devices	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem	09/12/2021	09/12/2021
									A070504	Premature Discharge of Battery	09/12/2021	09/12/2021

								A070909	Oversensing	09/12/2021	09/12/2021
						F08	Hospitalization or Prolonged Hospitalization	A0705	Battery Problem	09/12/2021	09/12/2021
								A070504	Premature Discharge of Battery	09/12/2021	09/12/2021
								A070909	Oversensing	09/12/2021	09/12/2021
						F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A0705	Battery Problem	09/12/2021	09/12/2021
								A070504	Premature Discharge of Battery	09/12/2021	09/12/2021
								A070909	Oversensing	09/12/2021	09/12/2021
						F1905	Device Revision or Replacement	A0705	Battery Problem	09/12/2021	09/12/2021
								A070504	Premature Discharge of Battery	09/12/2021	09/12/2021
								A070909	Oversensing	09/12/2021	09/12/2021
					CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem	09/12/2021	09/12/2021
									A070504	Premature Discharge of Battery	09/12/2021	09/12/2021
									A070909	Oversensing	09/12/2021	09/12/2021
							F08	Hospitalization or Prolonged Hospitalization	A0705	Battery Problem	09/12/2021	09/12/2021
									A070504	Premature Discharge of Battery	09/12/2021	09/12/2021
									A070909	Oversensing	09/12/2021	09/12/2021
							F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A0705	Battery Problem	09/12/2021	09/12/2021
									A070504	Premature Discharge of Battery	09/12/2021	09/12/2021
									A070909	Oversensing	09/12/2021	09/12/2021
							F1905	Device Revision or Replacement	A0705	Battery Problem	09/12/2021	09/12/2021
									A070504	Premature Discharge of Battery	09/12/2021	09/12/2021
									A070909	Oversensing	09/12/2021	09/12/2021
					CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem	09/12/2021	09/12/2021
									A070504	Premature Discharge of Battery	09/12/2021	09/12/2021
									A070909	Oversensing	09/12/2021	09/12/2021
							F08	Hospitalization or Prolonged Hospitalization	A0705	Battery Problem	09/12/2021	09/12/2021
									A070504	Premature Discharge of Battery	09/12/2021	09/12/2021
									A070909	Oversensing	09/12/2021	09/12/2021
							F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A0705	Battery Problem	09/12/2021	09/12/2021
									A070504	Premature Discharge of Battery	09/12/2021	09/12/2021
									A070909	Oversensing	09/12/2021	09/12/2021
							F1905	Device Revision or Replacement	A0705	Battery Problem	09/12/2021	09/12/2021
									A070504	Premature Discharge of Battery	09/12/2021	09/12/2021
									A070909	Oversensing	09/12/2021	09/12/2021
					CT233	Surgical	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem	09/12/2021	09/12/2021
									A070504	Premature Discharge of Battery	09/12/2021	09/12/2021
									A070909	Oversensing	09/12/2021	09/12/2021
							F08	Hospitalization or Prolonged Hospitalization	A0705	Battery Problem	09/12/2021	09/12/2021
									A070504	Premature Discharge of Battery	09/12/2021	09/12/2021
									A070909	Oversensing	09/12/2021	09/12/2021
							F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A0705	Battery Problem	09/12/2021	09/12/2021
									A070504	Premature Discharge of Battery	09/12/2021	09/12/2021
									A070909	Oversensing	09/12/2021	09/12/2021
							F1905	Device Revision or Replacement	A0705	Battery Problem	09/12/2021	09/12/2021
									A070504	Premature Discharge of Battery	09/12/2021	09/12/2021
									A070909	Oversensing	09/12/2021	09/12/2021
					CT2406	Implantable	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem	09/12/2021	09/12/2021
									A070504	Premature Discharge of Battery	09/12/2021	09/12/2021
									A070909	Oversensing	09/12/2021	09/12/2021
							F08	Hospitalization or Prolonged Hospitalization	A0705	Battery Problem	09/12/2021	09/12/2021
									A070504	Premature Discharge of Battery	09/12/2021	09/12/2021
									A070909	Oversensing	09/12/2021	09/12/2021
							F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A0705	Battery Problem	09/12/2021	09/12/2021
									A070504	Premature Discharge of Battery	09/12/2021	09/12/2021
									A070909	Oversensing	09/12/2021	09/12/2021
							F1905	Device Revision or Replacement	A0705	Battery Problem	09/12/2021	09/12/2021
									A070504	Premature Discharge of Battery	09/12/2021	09/12/2021
									A070909	Oversensing	09/12/2021	09/12/2021
					CT2407	Active implantable	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem	09/12/2021	09/12/2021
									A070504	Premature Discharge of Battery	09/12/2021	09/12/2021
									A070909	Oversensing	09/12/2021	09/12/2021
							F08	Hospitalization or Prolonged Hospitalization	A0705	Battery Problem	09/12/2021	09/12/2021
									A070504	Premature Discharge of Battery	09/12/2021	09/12/2021
									A070909	Oversensing	09/12/2021	09/12/2021
							F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A0705	Battery Problem	09/12/2021	09/12/2021
									A070504	Premature Discharge of Battery	09/12/2021	09/12/2021
									A070909	Oversensing	09/12/2021	09/12/2021
							F1905	Device Revision or Replacement	A0705	Battery Problem	09/12/2021	09/12/2021
									A070504	Premature Discharge of Battery	09/12/2021	09/12/2021
									A070909	Oversensing	09/12/2021	09/12/2021
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem	09/12/2021	09/12/2021
									A070504	Premature Discharge of Battery	09/12/2021	09/12/2021
									A070909	Oversensing	09/12/2021	09/12/2021
							F08	Hospitalization or Prolonged Hospitalization	A0705	Battery Problem	09/12/2021	09/12/2021
									A070504	Premature Discharge of Battery	09/12/2021	09/12/2021
									A070909	Oversensing	09/12/2021	09/12/2021
							F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A0705	Battery Problem	09/12/2021	09/12/2021
									A070504	Premature Discharge of Battery	09/12/2021	09/12/2021
									A070909	Oversensing	09/12/2021	09/12/2021
							F1905	Device Revision or Replacement	A0705	Battery Problem	09/12/2021	09/12/2021
									A070504	Premature Discharge of Battery	09/12/2021	09/12/2021
									A070909	Oversensing	09/12/2021	09/12/2021
					CT2727	Active implant devices/ancillaries	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem	09/12/2021	09/12/2021
									A070504	Premature Discharge of Battery	09/12/2021	09/12/2021
									A070909	Oversensing	09/12/2021	09/12/2021
							F08	Hospitalization or Prolonged Hospitalization	A0705	Battery Problem	09/12/2021	09/12/2021
									A070504	Premature Discharge of Battery	09/12/2021	09/12/2021
									A070909	Oversensing	09/12/2021	09/12/2021

							F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A0705	Battery Problem	09/12/2021	09/12/2021
									A070504	Premature Discharge of Battery	09/12/2021	09/12/2021
									A070909	Oversensing	09/12/2021	09/12/2021
							F1905	Device Revision or Replacement	A0705	Battery Problem	09/12/2021	09/12/2021
									A070504	Premature Discharge of Battery	09/12/2021	09/12/2021
									A070909	Oversensing	09/12/2021	09/12/2021
					CT2750	Implantable cardiac electrical stimulators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem	09/12/2021	09/12/2021
									A070504	Premature Discharge of Battery	09/12/2021	09/12/2021
									A070909	Oversensing	09/12/2021	09/12/2021
							F08	Hospitalization or Prolonged Hospitalization	A0705	Battery Problem	09/12/2021	09/12/2021
									A070504	Premature Discharge of Battery	09/12/2021	09/12/2021
									A070909	Oversensing	09/12/2021	09/12/2021
							F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A0705	Battery Problem	09/12/2021	09/12/2021
									A070504	Premature Discharge of Battery	09/12/2021	09/12/2021
									A070909	Oversensing	09/12/2021	09/12/2021
							F1905	Device Revision or Replacement	A0705	Battery Problem	09/12/2021	09/12/2021
									A070504	Premature Discharge of Battery	09/12/2021	09/12/2021
									A070909	Oversensing	09/12/2021	09/12/2021
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem	09/12/2021	09/12/2021
									A070504	Premature Discharge of Battery	09/12/2021	09/12/2021
									A070909	Oversensing	09/12/2021	09/12/2021
							F08	Hospitalization or Prolonged Hospitalization	A0705	Battery Problem	09/12/2021	09/12/2021
									A070504	Premature Discharge of Battery	09/12/2021	09/12/2021
									A070909	Oversensing	09/12/2021	09/12/2021
							F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A0705	Battery Problem	09/12/2021	09/12/2021
									A070504	Premature Discharge of Battery	09/12/2021	09/12/2021
									A070909	Oversensing	09/12/2021	09/12/2021
							F1905	Device Revision or Replacement	A0705	Battery Problem	09/12/2021	09/12/2021
									A070504	Premature Discharge of Battery	09/12/2021	09/12/2021
									A070909	Oversensing	09/12/2021	09/12/2021
					CT2985	Active	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem	09/12/2021	09/12/2021
									A070504	Premature Discharge of Battery	09/12/2021	09/12/2021
									A070909	Oversensing	09/12/2021	09/12/2021
							F08	Hospitalization or Prolonged Hospitalization	A0705	Battery Problem	09/12/2021	09/12/2021
									A070504	Premature Discharge of Battery	09/12/2021	09/12/2021
									A070909	Oversensing	09/12/2021	09/12/2021
							F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A0705	Battery Problem	09/12/2021	09/12/2021
									A070504	Premature Discharge of Battery	09/12/2021	09/12/2021
									A070909	Oversensing	09/12/2021	09/12/2021
							F1905	Device Revision or Replacement	A0705	Battery Problem	09/12/2021	09/12/2021
									A070504	Premature Discharge of Battery	09/12/2021	09/12/2021
									A070909	Oversensing	09/12/2021	09/12/2021
					CT3001	Implantable metal	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem	09/12/2021	09/12/2021
									A070504	Premature Discharge of Battery	09/12/2021	09/12/2021
									A070909	Oversensing	09/12/2021	09/12/2021
							F08	Hospitalization or Prolonged Hospitalization	A0705	Battery Problem	09/12/2021	09/12/2021
									A070504	Premature Discharge of Battery	09/12/2021	09/12/2021
									A070909	Oversensing	09/12/2021	09/12/2021
							F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A0705	Battery Problem	09/12/2021	09/12/2021
									A070504	Premature Discharge of Battery	09/12/2021	09/12/2021
									A070909	Oversensing	09/12/2021	09/12/2021
							F1905	Device Revision or Replacement	A0705	Battery Problem	09/12/2021	09/12/2021
									A070504	Premature Discharge of Battery	09/12/2021	09/12/2021
									A070909	Oversensing	09/12/2021	09/12/2021
					CT3002	Implantable synthetic polymer	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem	09/12/2021	09/12/2021
									A070504	Premature Discharge of Battery	09/12/2021	09/12/2021
									A070909	Oversensing	09/12/2021	09/12/2021
							F08	Hospitalization or Prolonged Hospitalization	A0705	Battery Problem	09/12/2021	09/12/2021
									A070504	Premature Discharge of Battery	09/12/2021	09/12/2021
									A070909	Oversensing	09/12/2021	09/12/2021
							F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A0705	Battery Problem	09/12/2021	09/12/2021
									A070504	Premature Discharge of Battery	09/12/2021	09/12/2021
									A070909	Oversensing	09/12/2021	09/12/2021
							F1905	Device Revision or Replacement	A0705	Battery Problem	09/12/2021	09/12/2021
									A070504	Premature Discharge of Battery	09/12/2021	09/12/2021
									A070909	Oversensing	09/12/2021	09/12/2021
					CT3006	Implantable inorganic material	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem	09/12/2021	09/12/2021
									A070504	Premature Discharge of Battery	09/12/2021	09/12/2021
									A070909	Oversensing	09/12/2021	09/12/2021
							F08	Hospitalization or Prolonged Hospitalization	A0705	Battery Problem	09/12/2021	09/12/2021
									A070504	Premature Discharge of Battery	09/12/2021	09/12/2021
									A070909	Oversensing	09/12/2021	09/12/2021
							F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A0705	Battery Problem	09/12/2021	09/12/2021
									A070504	Premature Discharge of Battery	09/12/2021	09/12/2021
									A070909	Oversensing	09/12/2021	09/12/2021
							F1905	Device Revision or Replacement	A0705	Battery Problem	09/12/2021	09/12/2021
									A070504	Premature Discharge of Battery	09/12/2021	09/12/2021
									A070909	Oversensing	09/12/2021	09/12/2021
					CT314	Electrically powered	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem	09/12/2021	09/12/2021
									A070504	Premature Discharge of Battery	09/12/2021	09/12/2021
									A070909	Oversensing	09/12/2021	09/12/2021
							F08	Hospitalization or Prolonged Hospitalization	A0705	Battery Problem	09/12/2021	09/12/2021
									A070504	Premature Discharge of Battery	09/12/2021	09/12/2021
									A070909	Oversensing	09/12/2021	09/12/2021
							F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A0705	Battery Problem	09/12/2021	09/12/2021
									A070504	Premature Discharge of Battery	09/12/2021	09/12/2021
									A070909	Oversensing	09/12/2021	09/12/2021
							F1905	Device Revision or Replacement	A0705	Battery Problem	09/12/2021	09/12/2021

									A070504	Premature Discharge of Battery	09/12/2021	09/12/2021
									A070909	Oversensing	09/12/2021	09/12/2021
					CT321	Long-term surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem	09/12/2021	09/12/2021
									A070504	Premature Discharge of Battery	09/12/2021	09/12/2021
									A070909	Oversensing	09/12/2021	09/12/2021
							F08	Hospitalization or Prolonged Hospitalization	A0705	Battery Problem	09/12/2021	09/12/2021
									A070504	Premature Discharge of Battery	09/12/2021	09/12/2021
									A070909	Oversensing	09/12/2021	09/12/2021
							F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A0705	Battery Problem	09/12/2021	09/12/2021
									A070504	Premature Discharge of Battery	09/12/2021	09/12/2021
									A070909	Oversensing	09/12/2021	09/12/2021
							F1905	Device Revision or Replacement	A0705	Battery Problem	09/12/2021	09/12/2021
									A070504	Premature Discharge of Battery	09/12/2021	09/12/2021
									A070909	Oversensing	09/12/2021	09/12/2021
					CT449	Generators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem	09/12/2021	09/12/2021
									A070504	Premature Discharge of Battery	09/12/2021	09/12/2021
									A070909	Oversensing	09/12/2021	09/12/2021
							F08	Hospitalization or Prolonged Hospitalization	A0705	Battery Problem	09/12/2021	09/12/2021
									A070504	Premature Discharge of Battery	09/12/2021	09/12/2021
									A070909	Oversensing	09/12/2021	09/12/2021
							F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A0705	Battery Problem	09/12/2021	09/12/2021
									A070504	Premature Discharge of Battery	09/12/2021	09/12/2021
									A070909	Oversensing	09/12/2021	09/12/2021
							F1905	Device Revision or Replacement	A0705	Battery Problem	09/12/2021	09/12/2021
									A070504	Premature Discharge of Battery	09/12/2021	09/12/2021
									A070909	Oversensing	09/12/2021	09/12/2021
					CT981	Single use	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem	09/12/2021	09/12/2021
									A070504	Premature Discharge of Battery	09/12/2021	09/12/2021
									A070909	Oversensing	09/12/2021	09/12/2021
							F08	Hospitalization or Prolonged Hospitalization	A0705	Battery Problem	09/12/2021	09/12/2021
									A070504	Premature Discharge of Battery	09/12/2021	09/12/2021
									A070909	Oversensing	09/12/2021	09/12/2021
							F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A0705	Battery Problem	09/12/2021	09/12/2021
									A070504	Premature Discharge of Battery	09/12/2021	09/12/2021
									A070909	Oversensing	09/12/2021	09/12/2021
							F1905	Device Revision or Replacement	A0705	Battery Problem	09/12/2021	09/12/2021
									A070504	Premature Discharge of Battery	09/12/2021	09/12/2021
									A070909	Oversensing	09/12/2021	09/12/2021
					CT983	Surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem	09/12/2021	09/12/2021
									A070504	Premature Discharge of Battery	09/12/2021	09/12/2021
									A070909	Oversensing	09/12/2021	09/12/2021
							F08	Hospitalization or Prolonged Hospitalization	A0705	Battery Problem	09/12/2021	09/12/2021
									A070504	Premature Discharge of Battery	09/12/2021	09/12/2021
									A070909	Oversensing	09/12/2021	09/12/2021
							F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A0705	Battery Problem	09/12/2021	09/12/2021
									A070504	Premature Discharge of Battery	09/12/2021	09/12/2021
									A070909	Oversensing	09/12/2021	09/12/2021
							F1905	Device Revision or Replacement	A0705	Battery Problem	09/12/2021	09/12/2021
									A070504	Premature Discharge of Battery	09/12/2021	09/12/2021
									A070909	Oversensing	09/12/2021	09/12/2021
7190	2021	All other reportable incidents	47267	Single-chamber implantable pacemaker, rate-responsive	CT141	Electrodes/Leads and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem	03/03/2022	03/03/2022
									A070504	Premature Discharge of Battery	03/03/2022	03/03/2022
							F08	Hospitalization or Prolonged Hospitalization	A0705	Battery Problem	03/03/2022	03/03/2022
									A070504	Premature Discharge of Battery	03/03/2022	03/03/2022
							F1905	Device Revision or Replacement	A0705	Battery Problem	03/03/2022	03/03/2022
									A070504	Premature Discharge of Battery	03/03/2022	03/03/2022
					CT145	Cardiovascular devices	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem	03/03/2022	03/03/2022
									A070504	Premature Discharge of Battery	03/03/2022	03/03/2022
							F08	Hospitalization or Prolonged Hospitalization	A0705	Battery Problem	03/03/2022	03/03/2022
									A070504	Premature Discharge of Battery	03/03/2022	03/03/2022
							F1905	Device Revision or Replacement	A0705	Battery Problem	03/03/2022	03/03/2022
									A070504	Premature Discharge of Battery	03/03/2022	03/03/2022
					CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem	03/03/2022	03/03/2022
									A070504	Premature Discharge of Battery	03/03/2022	03/03/2022
							F08	Hospitalization or Prolonged Hospitalization	A0705	Battery Problem	03/03/2022	03/03/2022
									A070504	Premature Discharge of Battery	03/03/2022	03/03/2022
							F1905	Device Revision or Replacement	A0705	Battery Problem	03/03/2022	03/03/2022
									A070504	Premature Discharge of Battery	03/03/2022	03/03/2022
					CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem	03/03/2022	03/03/2022
									A070504	Premature Discharge of Battery	03/03/2022	03/03/2022
							F08	Hospitalization or Prolonged Hospitalization	A0705	Battery Problem	03/03/2022	03/03/2022
									A070504	Premature Discharge of Battery	03/03/2022	03/03/2022
							F1905	Device Revision or Replacement	A0705	Battery Problem	03/03/2022	03/03/2022
									A070504	Premature Discharge of Battery	03/03/2022	03/03/2022
					CT233	Surgical	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem	03/03/2022	03/03/2022
									A070504	Premature Discharge of Battery	03/03/2022	03/03/2022
							F08	Hospitalization or Prolonged Hospitalization	A0705	Battery Problem	03/03/2022	03/03/2022
									A070504	Premature Discharge of Battery	03/03/2022	03/03/2022
							F1905	Device Revision or Replacement	A0705	Battery Problem	03/03/2022	03/03/2022
									A070504	Premature Discharge of Battery	03/03/2022	03/03/2022
					CT2406	Implantable	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem	03/03/2022	03/03/2022
									A070504	Premature Discharge of Battery	03/03/2022	03/03/2022
							F08	Hospitalization or Prolonged Hospitalization	A0705	Battery Problem	03/03/2022	03/03/2022
									A070504	Premature Discharge of Battery	03/03/2022	03/03/2022
							F1905	Device Revision or Replacement	A0705	Battery Problem	03/03/2022	03/03/2022
									A070504	Premature Discharge of Battery	03/03/2022	03/03/2022
					CT2407	Active implantable	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem	03/03/2022	03/03/2022
									A070504	Premature Discharge of Battery	03/03/2022	03/03/2022

							F08	Hospitalization or Prolonged Hospitalization	A0705	Battery Problem	03/03/2022	03/03/2022
									A070504	Premature Discharge of Battery	03/03/2022	03/03/2022
							F1905	Device Revision or Replacement	A0705	Battery Problem	03/03/2022	03/03/2022
									A070504	Premature Discharge of Battery	03/03/2022	03/03/2022
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem	03/03/2022	03/03/2022
									A070504	Premature Discharge of Battery	03/03/2022	03/03/2022
							F08	Hospitalization or Prolonged Hospitalization	A0705	Battery Problem	03/03/2022	03/03/2022
									A070504	Premature Discharge of Battery	03/03/2022	03/03/2022
							F1905	Device Revision or Replacement	A0705	Battery Problem	03/03/2022	03/03/2022
									A070504	Premature Discharge of Battery	03/03/2022	03/03/2022
					CT2727	Active implant devices/ancillaries	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem	03/03/2022	03/03/2022
									A070504	Premature Discharge of Battery	03/03/2022	03/03/2022
							F08	Hospitalization or Prolonged Hospitalization	A0705	Battery Problem	03/03/2022	03/03/2022
									A070504	Premature Discharge of Battery	03/03/2022	03/03/2022
							F1905	Device Revision or Replacement	A0705	Battery Problem	03/03/2022	03/03/2022
									A070504	Premature Discharge of Battery	03/03/2022	03/03/2022
					CT2750	Implantable cardiac electrical stimulators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem	03/03/2022	03/03/2022
									A070504	Premature Discharge of Battery	03/03/2022	03/03/2022
							F08	Hospitalization or Prolonged Hospitalization	A0705	Battery Problem	03/03/2022	03/03/2022
									A070504	Premature Discharge of Battery	03/03/2022	03/03/2022
							F1905	Device Revision or Replacement	A0705	Battery Problem	03/03/2022	03/03/2022
									A070504	Premature Discharge of Battery	03/03/2022	03/03/2022
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem	03/03/2022	03/03/2022
									A070504	Premature Discharge of Battery	03/03/2022	03/03/2022
							F08	Hospitalization or Prolonged Hospitalization	A0705	Battery Problem	03/03/2022	03/03/2022
									A070504	Premature Discharge of Battery	03/03/2022	03/03/2022
							F1905	Device Revision or Replacement	A0705	Battery Problem	03/03/2022	03/03/2022
									A070504	Premature Discharge of Battery	03/03/2022	03/03/2022
					CT2985	Active	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem	03/03/2022	03/03/2022
									A070504	Premature Discharge of Battery	03/03/2022	03/03/2022
							F08	Hospitalization or Prolonged Hospitalization	A0705	Battery Problem	03/03/2022	03/03/2022
									A070504	Premature Discharge of Battery	03/03/2022	03/03/2022
							F1905	Device Revision or Replacement	A0705	Battery Problem	03/03/2022	03/03/2022
									A070504	Premature Discharge of Battery	03/03/2022	03/03/2022
					CT3001	Implantable metal	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem	03/03/2022	03/03/2022
									A070504	Premature Discharge of Battery	03/03/2022	03/03/2022
							F08	Hospitalization or Prolonged Hospitalization	A0705	Battery Problem	03/03/2022	03/03/2022
									A070504	Premature Discharge of Battery	03/03/2022	03/03/2022
							F1905	Device Revision or Replacement	A0705	Battery Problem	03/03/2022	03/03/2022
									A070504	Premature Discharge of Battery	03/03/2022	03/03/2022
					CT3002	Implantable synthetic polymer	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem	03/03/2022	03/03/2022
									A070504	Premature Discharge of Battery	03/03/2022	03/03/2022
							F08	Hospitalization or Prolonged Hospitalization	A0705	Battery Problem	03/03/2022	03/03/2022
									A070504	Premature Discharge of Battery	03/03/2022	03/03/2022
							F1905	Device Revision or Replacement	A0705	Battery Problem	03/03/2022	03/03/2022
									A070504	Premature Discharge of Battery	03/03/2022	03/03/2022
					CT3006	Implantable inorganic material	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem	03/03/2022	03/03/2022
									A070504	Premature Discharge of Battery	03/03/2022	03/03/2022
							F08	Hospitalization or Prolonged Hospitalization	A0705	Battery Problem	03/03/2022	03/03/2022
									A070504	Premature Discharge of Battery	03/03/2022	03/03/2022
							F1905	Device Revision or Replacement	A0705	Battery Problem	03/03/2022	03/03/2022
									A070504	Premature Discharge of Battery	03/03/2022	03/03/2022
					CT314	Electrically powered	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem	03/03/2022	03/03/2022
									A070504	Premature Discharge of Battery	03/03/2022	03/03/2022
							F08	Hospitalization or Prolonged Hospitalization	A0705	Battery Problem	03/03/2022	03/03/2022
									A070504	Premature Discharge of Battery	03/03/2022	03/03/2022
							F1905	Device Revision or Replacement	A0705	Battery Problem	03/03/2022	03/03/2022
									A070504	Premature Discharge of Battery	03/03/2022	03/03/2022
					CT321	Long-term surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem	03/03/2022	03/03/2022
									A070504	Premature Discharge of Battery	03/03/2022	03/03/2022
							F08	Hospitalization or Prolonged Hospitalization	A0705	Battery Problem	03/03/2022	03/03/2022
									A070504	Premature Discharge of Battery	03/03/2022	03/03/2022
							F1905	Device Revision or Replacement	A0705	Battery Problem	03/03/2022	03/03/2022
									A070504	Premature Discharge of Battery	03/03/2022	03/03/2022
					CT449	Generators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem	03/03/2022	03/03/2022
									A070504	Premature Discharge of Battery	03/03/2022	03/03/2022
							F08	Hospitalization or Prolonged Hospitalization	A0705	Battery Problem	03/03/2022	03/03/2022
									A070504	Premature Discharge of Battery	03/03/2022	03/03/2022
							F1905	Device Revision or Replacement	A0705	Battery Problem	03/03/2022	03/03/2022
									A070504	Premature Discharge of Battery	03/03/2022	03/03/2022
					CT981	Single use	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem	03/03/2022	03/03/2022
									A070504	Premature Discharge of Battery	03/03/2022	03/03/2022
							F08	Hospitalization or Prolonged Hospitalization	A0705	Battery Problem	03/03/2022	03/03/2022
									A070504	Premature Discharge of Battery	03/03/2022	03/03/2022
							F1905	Device Revision or Replacement	A0705	Battery Problem	03/03/2022	03/03/2022
									A070504	Premature Discharge of Battery	03/03/2022	03/03/2022
					CT983	Surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem	03/03/2022	03/03/2022
									A070504	Premature Discharge of Battery	03/03/2022	03/03/2022
							F08	Hospitalization or Prolonged Hospitalization	A0705	Battery Problem	03/03/2022	03/03/2022
									A070504	Premature Discharge of Battery	03/03/2022	03/03/2022
							F1905	Device Revision or Replacement	A0705	Battery Problem	03/03/2022	03/03/2022
									A070504	Premature Discharge of Battery	03/03/2022	03/03/2022
7191	2021	Death	47270	Cardiac resynchronization therapy implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E060109	Tachycardia	A0709	Device Sensing Problem	20/11/2021	20/11/2021
							E060110	Ventricular Fibrillation	A0709	Device Sensing Problem	20/11/2021	20/11/2021
							E0602	Cardiac Arrest	A0709	Device Sensing Problem	20/11/2021	20/11/2021
							F02	Death	A0709	Device Sensing Problem	20/11/2021	20/11/2021
							F1203	Life Threatening Illness or Injury	A0709	Device Sensing Problem	20/11/2021	20/11/2021
							F2306	Resuscitation	A0709	Device Sensing Problem	20/11/2021	20/11/2021

7192	2021	All other reportable incidents	17882	Physiologic-monitoring defibrillation system	CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A191002	Loss of Power		
							F26	No Health Consequences or Impact	A191002	Loss of Power		
7193	2021	All other reportable incidents	17882	Physiologic-monitoring defibrillation system	CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A071301	Failure to Deliver Shock/Stimulation		
							F26	No Health Consequences or Impact	A071301	Failure to Deliver Shock/Stimulation		
7194	2022	All other reportable incidents	17882	Physiologic-monitoring defibrillation system	CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A191002	Loss of Power		
							F26	No Health Consequences or Impact	A191002	Loss of Power		
7195	2021	Death	17882	Physiologic-monitoring defibrillation system	CT1684	Cardiac pacemakers and associated devices	E0602	Cardiac Arrest	A071301	Failure to Deliver Shock/Stimulation		
							F02	Death	A071301	Failure to Deliver Shock/Stimulation		
7196	2022	All other reportable incidents	47267	Single-chamber implantable pacemaker, rate-responsive	CT141	Electrodes/Leads and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem	10/10/2022	10/10/2022
									A070504	Premature Discharge of Battery	10/10/2022	10/10/2022
							F08	Hospitalization or Prolonged Hospitalization	A0705	Battery Problem	10/10/2022	10/10/2022
									A070504	Premature Discharge of Battery	10/10/2022	10/10/2022
							F1905	Device Revision or Replacement	A0705	Battery Problem	10/10/2022	10/10/2022
									A070504	Premature Discharge of Battery	10/10/2022	10/10/2022
					CT145	Cardiovascular devices	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem	10/10/2022	10/10/2022
									A070504	Premature Discharge of Battery	10/10/2022	10/10/2022
							F08	Hospitalization or Prolonged Hospitalization	A0705	Battery Problem	10/10/2022	10/10/2022
									A070504	Premature Discharge of Battery	10/10/2022	10/10/2022
							F1905	Device Revision or Replacement	A0705	Battery Problem	10/10/2022	10/10/2022
									A070504	Premature Discharge of Battery	10/10/2022	10/10/2022
					CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem	10/10/2022	10/10/2022
									A070504	Premature Discharge of Battery	10/10/2022	10/10/2022
							F08	Hospitalization or Prolonged Hospitalization	A0705	Battery Problem	10/10/2022	10/10/2022
									A070504	Premature Discharge of Battery	10/10/2022	10/10/2022
							F1905	Device Revision or Replacement	A0705	Battery Problem	10/10/2022	10/10/2022
									A070504	Premature Discharge of Battery	10/10/2022	10/10/2022
					CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem	10/10/2022	10/10/2022
									A070504	Premature Discharge of Battery	10/10/2022	10/10/2022
							F08	Hospitalization or Prolonged Hospitalization	A0705	Battery Problem	10/10/2022	10/10/2022
									A070504	Premature Discharge of Battery	10/10/2022	10/10/2022
							F1905	Device Revision or Replacement	A0705	Battery Problem	10/10/2022	10/10/2022
									A070504	Premature Discharge of Battery	10/10/2022	10/10/2022
					CT233	Surgical	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem	10/10/2022	10/10/2022
									A070504	Premature Discharge of Battery	10/10/2022	10/10/2022
							F08	Hospitalization or Prolonged Hospitalization	A0705	Battery Problem	10/10/2022	10/10/2022
									A070504	Premature Discharge of Battery	10/10/2022	10/10/2022
							F1905	Device Revision or Replacement	A0705	Battery Problem	10/10/2022	10/10/2022
									A070504	Premature Discharge of Battery	10/10/2022	10/10/2022
					CT2406	Implantable	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem	10/10/2022	10/10/2022
									A070504	Premature Discharge of Battery	10/10/2022	10/10/2022
							F08	Hospitalization or Prolonged Hospitalization	A0705	Battery Problem	10/10/2022	10/10/2022
									A070504	Premature Discharge of Battery	10/10/2022	10/10/2022
							F1905	Device Revision or Replacement	A0705	Battery Problem	10/10/2022	10/10/2022
									A070504	Premature Discharge of Battery	10/10/2022	10/10/2022
					CT2407	Active implantable	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem	10/10/2022	10/10/2022
									A070504	Premature Discharge of Battery	10/10/2022	10/10/2022
							F08	Hospitalization or Prolonged Hospitalization	A0705	Battery Problem	10/10/2022	10/10/2022
									A070504	Premature Discharge of Battery	10/10/2022	10/10/2022
							F1905	Device Revision or Replacement	A0705	Battery Problem	10/10/2022	10/10/2022
									A070504	Premature Discharge of Battery	10/10/2022	10/10/2022
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem	10/10/2022	10/10/2022
									A070504	Premature Discharge of Battery	10/10/2022	10/10/2022
							F08	Hospitalization or Prolonged Hospitalization	A0705	Battery Problem	10/10/2022	10/10/2022
									A070504	Premature Discharge of Battery	10/10/2022	10/10/2022
							F1905	Device Revision or Replacement	A0705	Battery Problem	10/10/2022	10/10/2022
									A070504	Premature Discharge of Battery	10/10/2022	10/10/2022
					CT2727	Active implant devices/ancillaries	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem	10/10/2022	10/10/2022
									A070504	Premature Discharge of Battery	10/10/2022	10/10/2022
							F08	Hospitalization or Prolonged Hospitalization	A0705	Battery Problem	10/10/2022	10/10/2022
									A070504	Premature Discharge of Battery	10/10/2022	10/10/2022
							F1905	Device Revision or Replacement	A0705	Battery Problem	10/10/2022	10/10/2022
									A070504	Premature Discharge of Battery	10/10/2022	10/10/2022
					CT2750	Implantable cardiac electrical stimulators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem	10/10/2022	10/10/2022
									A070504	Premature Discharge of Battery	10/10/2022	10/10/2022
							F08	Hospitalization or Prolonged Hospitalization	A0705	Battery Problem	10/10/2022	10/10/2022
									A070504	Premature Discharge of Battery	10/10/2022	10/10/2022
							F1905	Device Revision or Replacement	A0705	Battery Problem	10/10/2022	10/10/2022
									A070504	Premature Discharge of Battery	10/10/2022	10/10/2022
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem	10/10/2022	10/10/2022
									A070504	Premature Discharge of Battery	10/10/2022	10/10/2022
							F08	Hospitalization or Prolonged Hospitalization	A0705	Battery Problem	10/10/2022	10/10/2022
									A070504	Premature Discharge of Battery	10/10/2022	10/10/2022
							F1905	Device Revision or Replacement	A0705	Battery Problem	10/10/2022	10/10/2022
									A070504	Premature Discharge of Battery	10/10/2022	10/10/2022
					CT2985	Active	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem	10/10/2022	10/10/2022
									A070504	Premature Discharge of Battery	10/10/2022	10/10/2022
							F08	Hospitalization or Prolonged Hospitalization	A0705	Battery Problem	10/10/2022	10/10/2022
									A070504	Premature Discharge of Battery	10/10/2022	10/10/2022
							F1905	Device Revision or Replacement	A0705	Battery Problem	10/10/2022	10/10/2022
									A070504	Premature Discharge of Battery	10/10/2022	10/10/2022
					CT3001	Implantable metal	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem	10/10/2022	10/10/2022
									A070504	Premature Discharge of Battery	10/10/2022	10/10/2022
							F08	Hospitalization or Prolonged Hospitalization	A0705	Battery Problem	10/10/2022	10/10/2022
									A070504	Premature Discharge of Battery	10/10/2022	10/10/2022
							F1905	Device Revision or Replacement	A0705	Battery Problem	10/10/2022	10/10/2022
									A070504	Premature Discharge of Battery	10/10/2022	10/10/2022
					CT3002	Implantable synthetic polymer	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem	10/10/2022	10/10/2022
									A070504	Premature Discharge of Battery	10/10/2022	10/10/2022

							F08	Hospitalization or Prolonged Hospitalization	A0705	Battery Problem	10/10/2022	10/10/2022
									A070504	Premature Discharge of Battery	10/10/2022	10/10/2022
							F1905	Device Revision or Replacement	A0705	Battery Problem	10/10/2022	10/10/2022
									A070504	Premature Discharge of Battery	10/10/2022	10/10/2022
					CT3006	Implantable inorganic material	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem	10/10/2022	10/10/2022
									A070504	Premature Discharge of Battery	10/10/2022	10/10/2022
							F08	Hospitalization or Prolonged Hospitalization	A0705	Battery Problem	10/10/2022	10/10/2022
									A070504	Premature Discharge of Battery	10/10/2022	10/10/2022
							F1905	Device Revision or Replacement	A0705	Battery Problem	10/10/2022	10/10/2022
									A070504	Premature Discharge of Battery	10/10/2022	10/10/2022
					CT314	Electrically powered	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem	10/10/2022	10/10/2022
									A070504	Premature Discharge of Battery	10/10/2022	10/10/2022
							F08	Hospitalization or Prolonged Hospitalization	A0705	Battery Problem	10/10/2022	10/10/2022
									A070504	Premature Discharge of Battery	10/10/2022	10/10/2022
							F1905	Device Revision or Replacement	A0705	Battery Problem	10/10/2022	10/10/2022
									A070504	Premature Discharge of Battery	10/10/2022	10/10/2022
					CT321	Long-term surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem	10/10/2022	10/10/2022
									A070504	Premature Discharge of Battery	10/10/2022	10/10/2022
							F08	Hospitalization or Prolonged Hospitalization	A0705	Battery Problem	10/10/2022	10/10/2022
									A070504	Premature Discharge of Battery	10/10/2022	10/10/2022
							F1905	Device Revision or Replacement	A0705	Battery Problem	10/10/2022	10/10/2022
									A070504	Premature Discharge of Battery	10/10/2022	10/10/2022
					CT449	Generators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem	10/10/2022	10/10/2022
									A070504	Premature Discharge of Battery	10/10/2022	10/10/2022
							F08	Hospitalization or Prolonged Hospitalization	A0705	Battery Problem	10/10/2022	10/10/2022
									A070504	Premature Discharge of Battery	10/10/2022	10/10/2022
							F1905	Device Revision or Replacement	A0705	Battery Problem	10/10/2022	10/10/2022
									A070504	Premature Discharge of Battery	10/10/2022	10/10/2022
					CT981	Single use	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem	10/10/2022	10/10/2022
									A070504	Premature Discharge of Battery	10/10/2022	10/10/2022
							F08	Hospitalization or Prolonged Hospitalization	A0705	Battery Problem	10/10/2022	10/10/2022
									A070504	Premature Discharge of Battery	10/10/2022	10/10/2022
							F1905	Device Revision or Replacement	A0705	Battery Problem	10/10/2022	10/10/2022
									A070504	Premature Discharge of Battery	10/10/2022	10/10/2022
					CT983	Surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem	10/10/2022	10/10/2022
									A070504	Premature Discharge of Battery	10/10/2022	10/10/2022
							F08	Hospitalization or Prolonged Hospitalization	A0705	Battery Problem	10/10/2022	10/10/2022
									A070504	Premature Discharge of Battery	10/10/2022	10/10/2022
							F1905	Device Revision or Replacement	A0705	Battery Problem	10/10/2022	10/10/2022
									A070504	Premature Discharge of Battery	10/10/2022	10/10/2022
7197	2022	All other reportable incidents	37265	Dual-chamber implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
							F24	Insufficient Health Impact Information	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
7198	2021	All other reportable incidents	17882	Physiologic-monitoring defibrillation system	CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A07	Electrical /Electronic Property Problem		
							F27	Problem Identified During Non-Clinical Procedure	A07	Electrical /Electronic Property Problem		
7199	2021	All other reportable incidents	35852	Single-chamber implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem	13/12/2021	13/12/2021
									A070504	Premature Discharge of Battery	13/12/2021	13/12/2021
							F08	Hospitalization or Prolonged Hospitalization	A0705	Battery Problem	13/12/2021	13/12/2021
									A070504	Premature Discharge of Battery	13/12/2021	13/12/2021
							F1903	Device Explantation	A0705	Battery Problem	13/12/2021	13/12/2021
									A070504	Premature Discharge of Battery	13/12/2021	13/12/2021
7200	2022	All other reportable incidents	36241	Pacemaker/defibrillator lead	CT1684	Cardiac pacemakers and associated devices	E2104	Electric Shock	A051201	Device Dislodged or Dislocated		
									A071302	Inappropriate/Inadequate Shock/Stimulation		
									A072201	High Impedance		
							E2403	No Clinical Signs, Symptoms or Conditions	A051201	Device Dislodged or Dislocated		
									A071302	Inappropriate/Inadequate Shock/Stimulation		
									A072201	High Impedance		
							F08	Hospitalization or Prolonged Hospitalization	A051201	Device Dislodged or Dislocated		
									A071302	Inappropriate/Inadequate Shock/Stimulation		
									A072201	High Impedance		
							F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A051201	Device Dislodged or Dislocated		
									A071302	Inappropriate/Inadequate Shock/Stimulation		
									A072201	High Impedance		
							F1905	Device Revision or Replacement	A051201	Device Dislodged or Dislocated		
									A071302	Inappropriate/Inadequate Shock/Stimulation		
									A072201	High Impedance		
7201	2022	All other reportable incidents	35852	Single-chamber implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E2104	Electric Shock	A010402	Migration		
									A071302	Inappropriate/Inadequate Shock/Stimulation		
									A072201	High Impedance		
							E2403	No Clinical Signs, Symptoms or Conditions	A010402	Migration		
									A071302	Inappropriate/Inadequate Shock/Stimulation		
									A072201	High Impedance		
							F08	Hospitalization or Prolonged Hospitalization	A010402	Migration		
									A071302	Inappropriate/Inadequate Shock/Stimulation		
									A072201	High Impedance		
							F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A010402	Migration		
									A071302	Inappropriate/Inadequate Shock/Stimulation		
									A072201	High Impedance		
							F1905	Device Revision or Replacement	A010402	Migration		
									A071302	Inappropriate/Inadequate Shock/Stimulation		
									A072201	High Impedance		
7202	2021	All other reportable incidents	35852	Single-chamber implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem	09/12/2021	09/12/2021
									A070504	Premature Discharge of Battery	09/12/2021	09/12/2021
							F08	Hospitalization or Prolonged Hospitalization	A0705	Battery Problem	09/12/2021	09/12/2021
									A070504	Premature Discharge of Battery	09/12/2021	09/12/2021
							F1905	Device Revision or Replacement	A0705	Battery Problem	09/12/2021	09/12/2021
									A070504	Premature Discharge of Battery	09/12/2021	09/12/2021

7203	2021	All other reportable incidents	35852	Single-chamber implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A0709	Device Sensing Problem		
									A1107	Data Problem		
							F26	No Health Consequences or Impact	A0709	Device Sensing Problem		
									A1107	Data Problem		
7204	2021	All other reportable incidents	36241	Pacemaker/defibrillator lead	CT1684	Cardiac pacemakers and associated devices	E0604	Cardiac Perforation	A24	No Apparent Device Problem		
							E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem		
							F08	Hospitalization or Prolonged Hospitalization	A24	No Apparent Device Problem		
							F1203	Life Threatening Illness or Injury	A24	No Apparent Device Problem		
							F1905	Device Revision or Replacement	A24	No Apparent Device Problem		
							F2203	Imaging Required	A24	No Apparent Device Problem		
7205	2021	All other reportable incidents	47270	Cardiac resynchronization therapy implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	08/10/2021	08/10/2021
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	08/10/2021	08/10/2021
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	08/10/2021	08/10/2021
7206	2022	All other reportable incidents	47270	Cardiac resynchronization therapy implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	07/07/2022	07/07/2022
							F1901	Additional Surgery	A070504	Premature Discharge of Battery	07/07/2022	07/07/2022
7207	2022	All other reportable incidents	17882	Physiologic-monitoring defibrillation system	CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem		
							F27	Problem Identified During Non-Clinical Procedure	A0705	Battery Problem		
7208	2022	All other reportable incidents	60789	Intracardiac pacemaker	CT130	Catheters and associated devices	E0604	Cardiac Perforation	A24	No Apparent Device Problem	08/07/2022	08/07/2022
							E0619	Pericardial Effusion	A24	No Apparent Device Problem	08/07/2022	08/07/2022
							E2321	Low Blood Pressure/ Hypotension	A24	No Apparent Device Problem	08/07/2022	08/07/2022
							E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	08/07/2022	08/07/2022
							F1203	Life Threatening Illness or Injury	A24	No Apparent Device Problem	08/07/2022	08/07/2022
							F19	Surgical Intervention	A24	No Apparent Device Problem	08/07/2022	08/07/2022
					CT145	Cardiovascular devices	E0604	Cardiac Perforation	A24	No Apparent Device Problem	08/07/2022	08/07/2022
							E0619	Pericardial Effusion	A24	No Apparent Device Problem	08/07/2022	08/07/2022
							E2321	Low Blood Pressure/ Hypotension	A24	No Apparent Device Problem	08/07/2022	08/07/2022
							E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	08/07/2022	08/07/2022
							F1203	Life Threatening Illness or Injury	A24	No Apparent Device Problem	08/07/2022	08/07/2022
							F19	Surgical Intervention	A24	No Apparent Device Problem	08/07/2022	08/07/2022
					CT1684	Cardiac pacemakers and associated devices	E0604	Cardiac Perforation	A24	No Apparent Device Problem	08/07/2022	08/07/2022
							E0619	Pericardial Effusion	A24	No Apparent Device Problem	08/07/2022	08/07/2022
							E2321	Low Blood Pressure/ Hypotension	A24	No Apparent Device Problem	08/07/2022	08/07/2022
							E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	08/07/2022	08/07/2022
							F1203	Life Threatening Illness or Injury	A24	No Apparent Device Problem	08/07/2022	08/07/2022
							F19	Surgical Intervention	A24	No Apparent Device Problem	08/07/2022	08/07/2022
					CT1685	Cardiac pacemakers	E0604	Cardiac Perforation	A24	No Apparent Device Problem	08/07/2022	08/07/2022
							E0619	Pericardial Effusion	A24	No Apparent Device Problem	08/07/2022	08/07/2022
							E2321	Low Blood Pressure/ Hypotension	A24	No Apparent Device Problem	08/07/2022	08/07/2022
							E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	08/07/2022	08/07/2022
							F1203	Life Threatening Illness or Injury	A24	No Apparent Device Problem	08/07/2022	08/07/2022
							F19	Surgical Intervention	A24	No Apparent Device Problem	08/07/2022	08/07/2022
					CT233	Surgical	E0604	Cardiac Perforation	A24	No Apparent Device Problem	08/07/2022	08/07/2022
							E0619	Pericardial Effusion	A24	No Apparent Device Problem	08/07/2022	08/07/2022
							E2321	Low Blood Pressure/ Hypotension	A24	No Apparent Device Problem	08/07/2022	08/07/2022
							E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	08/07/2022	08/07/2022
							F1203	Life Threatening Illness or Injury	A24	No Apparent Device Problem	08/07/2022	08/07/2022
							F19	Surgical Intervention	A24	No Apparent Device Problem	08/07/2022	08/07/2022
					CT2406	Implantable	E0604	Cardiac Perforation	A24	No Apparent Device Problem	08/07/2022	08/07/2022
							E0619	Pericardial Effusion	A24	No Apparent Device Problem	08/07/2022	08/07/2022
							E2321	Low Blood Pressure/ Hypotension	A24	No Apparent Device Problem	08/07/2022	08/07/2022
							E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	08/07/2022	08/07/2022
							F1203	Life Threatening Illness or Injury	A24	No Apparent Device Problem	08/07/2022	08/07/2022
							F19	Surgical Intervention	A24	No Apparent Device Problem	08/07/2022	08/07/2022
					CT2407	Active implantable	E0604	Cardiac Perforation	A24	No Apparent Device Problem	08/07/2022	08/07/2022
							E0619	Pericardial Effusion	A24	No Apparent Device Problem	08/07/2022	08/07/2022
							E2321	Low Blood Pressure/ Hypotension	A24	No Apparent Device Problem	08/07/2022	08/07/2022
							E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	08/07/2022	08/07/2022
							F1203	Life Threatening Illness or Injury	A24	No Apparent Device Problem	08/07/2022	08/07/2022
							F19	Surgical Intervention	A24	No Apparent Device Problem	08/07/2022	08/07/2022
					CT257	Implantable cardiac pacemakers	E0604	Cardiac Perforation	A24	No Apparent Device Problem	08/07/2022	08/07/2022
							E0619	Pericardial Effusion	A24	No Apparent Device Problem	08/07/2022	08/07/2022
							E2321	Low Blood Pressure/ Hypotension	A24	No Apparent Device Problem	08/07/2022	08/07/2022
							E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	08/07/2022	08/07/2022
							F1203	Life Threatening Illness or Injury	A24	No Apparent Device Problem	08/07/2022	08/07/2022
							F19	Surgical Intervention	A24	No Apparent Device Problem	08/07/2022	08/07/2022
					CT2727	Active implant devices/ancillaries	E0604	Cardiac Perforation	A24	No Apparent Device Problem	08/07/2022	08/07/2022
							E0619	Pericardial Effusion	A24	No Apparent Device Problem	08/07/2022	08/07/2022
							E2321	Low Blood Pressure/ Hypotension	A24	No Apparent Device Problem	08/07/2022	08/07/2022
							E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	08/07/2022	08/07/2022
							F1203	Life Threatening Illness or Injury	A24	No Apparent Device Problem	08/07/2022	08/07/2022
							F19	Surgical Intervention	A24	No Apparent Device Problem	08/07/2022	08/07/2022
					CT2750	Implantable cardiac electrical stimulators and associated devices	E0604	Cardiac Perforation	A24	No Apparent Device Problem	08/07/2022	08/07/2022
							E0619	Pericardial Effusion	A24	No Apparent Device Problem	08/07/2022	08/07/2022
							E2321	Low Blood Pressure/ Hypotension	A24	No Apparent Device Problem	08/07/2022	08/07/2022
							E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	08/07/2022	08/07/2022
							F1203	Life Threatening Illness or Injury	A24	No Apparent Device Problem	08/07/2022	08/07/2022
							F19	Surgical Intervention	A24	No Apparent Device Problem	08/07/2022	08/07/2022
					CT2818	Catheterization/Cannulation/Fluid path devices	E0604	Cardiac Perforation	A24	No Apparent Device Problem	08/07/2022	08/07/2022
							E0619	Pericardial Effusion	A24	No Apparent Device Problem	08/07/2022	08/07/2022
							E2321	Low Blood Pressure/ Hypotension	A24	No Apparent Device Problem	08/07/2022	08/07/2022
							E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	08/07/2022	08/07/2022
							F1203	Life Threatening Illness or Injury	A24	No Apparent Device Problem	08/07/2022	08/07/2022
							F19	Surgical Intervention	A24	No Apparent Device Problem	08/07/2022	08/07/2022
					CT2866	Implantable cardiac electrical stimulators	E0604	Cardiac Perforation	A24	No Apparent Device Problem	08/07/2022	08/07/2022
							E0619	Pericardial Effusion	A24	No Apparent Device Problem	08/07/2022	08/07/2022
							E2321	Low Blood Pressure/ Hypotension	A24	No Apparent Device Problem	08/07/2022	08/07/2022
							E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	08/07/2022	08/07/2022
							F1203	Life Threatening Illness or Injury	A24	No Apparent Device Problem	08/07/2022	08/07/2022

							F19	Surgical Intervention	A24	No Apparent Device Problem	08/07/2022	08/07/2022
					CT2985	Active	E0604	Cardiac Perforation	A24	No Apparent Device Problem	08/07/2022	08/07/2022
							E0619	Pericardial Effusion	A24	No Apparent Device Problem	08/07/2022	08/07/2022
							E2321	Low Blood Pressure/ Hypotension	A24	No Apparent Device Problem	08/07/2022	08/07/2022
							E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	08/07/2022	08/07/2022
							F1203	Life Threatening Illness or Injury	A24	No Apparent Device Problem	08/07/2022	08/07/2022
							F19	Surgical Intervention	A24	No Apparent Device Problem	08/07/2022	08/07/2022
					CT3001	Implantable metal	E0604	Cardiac Perforation	A24	No Apparent Device Problem	08/07/2022	08/07/2022
							E0619	Pericardial Effusion	A24	No Apparent Device Problem	08/07/2022	08/07/2022
							E2321	Low Blood Pressure/ Hypotension	A24	No Apparent Device Problem	08/07/2022	08/07/2022
							E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	08/07/2022	08/07/2022
							F1203	Life Threatening Illness or Injury	A24	No Apparent Device Problem	08/07/2022	08/07/2022
							F19	Surgical Intervention	A24	No Apparent Device Problem	08/07/2022	08/07/2022
					CT3002	Implantable synthetic polymer	E0604	Cardiac Perforation	A24	No Apparent Device Problem	08/07/2022	08/07/2022
							E0619	Pericardial Effusion	A24	No Apparent Device Problem	08/07/2022	08/07/2022
							E2321	Low Blood Pressure/ Hypotension	A24	No Apparent Device Problem	08/07/2022	08/07/2022
							E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	08/07/2022	08/07/2022
							F1203	Life Threatening Illness or Injury	A24	No Apparent Device Problem	08/07/2022	08/07/2022
							F19	Surgical Intervention	A24	No Apparent Device Problem	08/07/2022	08/07/2022
					CT3006	Implantable inorganic material	E0604	Cardiac Perforation	A24	No Apparent Device Problem	08/07/2022	08/07/2022
							E0619	Pericardial Effusion	A24	No Apparent Device Problem	08/07/2022	08/07/2022
							E2321	Low Blood Pressure/ Hypotension	A24	No Apparent Device Problem	08/07/2022	08/07/2022
							E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	08/07/2022	08/07/2022
							F1203	Life Threatening Illness or Injury	A24	No Apparent Device Problem	08/07/2022	08/07/2022
							F19	Surgical Intervention	A24	No Apparent Device Problem	08/07/2022	08/07/2022
					CT314	Electrically powered	E0604	Cardiac Perforation	A24	No Apparent Device Problem	08/07/2022	08/07/2022
							E0619	Pericardial Effusion	A24	No Apparent Device Problem	08/07/2022	08/07/2022
							E2321	Low Blood Pressure/ Hypotension	A24	No Apparent Device Problem	08/07/2022	08/07/2022
							E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	08/07/2022	08/07/2022
							F1203	Life Threatening Illness or Injury	A24	No Apparent Device Problem	08/07/2022	08/07/2022
							F19	Surgical Intervention	A24	No Apparent Device Problem	08/07/2022	08/07/2022
					CT321	Long-term surgical invasive	E0604	Cardiac Perforation	A24	No Apparent Device Problem	08/07/2022	08/07/2022
							E0619	Pericardial Effusion	A24	No Apparent Device Problem	08/07/2022	08/07/2022
							E2321	Low Blood Pressure/ Hypotension	A24	No Apparent Device Problem	08/07/2022	08/07/2022
							E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	08/07/2022	08/07/2022
							F1203	Life Threatening Illness or Injury	A24	No Apparent Device Problem	08/07/2022	08/07/2022
							F19	Surgical Intervention	A24	No Apparent Device Problem	08/07/2022	08/07/2022
					CT981	Single use	E0604	Cardiac Perforation	A24	No Apparent Device Problem	08/07/2022	08/07/2022
							E0619	Pericardial Effusion	A24	No Apparent Device Problem	08/07/2022	08/07/2022
							E2321	Low Blood Pressure/ Hypotension	A24	No Apparent Device Problem	08/07/2022	08/07/2022
							E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	08/07/2022	08/07/2022
							F1203	Life Threatening Illness or Injury	A24	No Apparent Device Problem	08/07/2022	08/07/2022
							F19	Surgical Intervention	A24	No Apparent Device Problem	08/07/2022	08/07/2022
					CT983	Surgical invasive	E0604	Cardiac Perforation	A24	No Apparent Device Problem	08/07/2022	08/07/2022
							E0619	Pericardial Effusion	A24	No Apparent Device Problem	08/07/2022	08/07/2022
							E2321	Low Blood Pressure/ Hypotension	A24	No Apparent Device Problem	08/07/2022	08/07/2022
							E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	08/07/2022	08/07/2022
							F1203	Life Threatening Illness or Injury	A24	No Apparent Device Problem	08/07/2022	08/07/2022
							F19	Surgical Intervention	A24	No Apparent Device Problem	08/07/2022	08/07/2022
7209	2022	All other reportable incidents	35223	Endocardial/interventricular septal pacing lead	CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070909	Oversensing		
									A072001	Electromagnetic Interference		
									A090801	Signal Artifact/Noise		
							F12	Serious Injury/ Illness/ Impairment	A070909	Oversensing		
									A072001	Electromagnetic Interference		
									A090801	Signal Artifact/Noise		
							F23	Unexpected Medical Intervention	A070909	Oversensing		
									A072001	Electromagnetic Interference		
									A090801	Signal Artifact/Noise		
7210	2022	All other reportable incidents	35223	Endocardial/interventricular septal pacing lead	CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A040101	Fracture		
									A070909	Oversensing		
									A072202	Low Impedance		
									A090801	Signal Artifact/Noise		
							F26	No Health Consequences or Impact	A040101	Fracture		
									A070909	Oversensing		
									A072202	Low Impedance		
									A090801	Signal Artifact/Noise		
7211	2021	All other reportable incidents	17882	Physiologic-monitoring defibrillation system	CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A1603	Failure of Device to Self-Test		
							F27	Problem Identified During Non-Clinical Procedure	A1603	Failure of Device to Self-Test		
7212	2021	All other reportable incidents	17882	Physiologic-monitoring defibrillation system	CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A07	Electrical /Electronic Property Problem		
							F27	Problem Identified During Non-Clinical Procedure	A07	Electrical /Electronic Property Problem		
7213	2021	All other reportable incidents	35223	Endocardial/interventricular septal pacing lead	CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070909	Oversensing		
							F26	No Health Consequences or Impact	A070909	Oversensing		
7214	2021	All other reportable incidents	35223	Endocardial/interventricular septal pacing lead	CT1684	Cardiac pacemakers and associated devices	E011903	Syncope/Fainting	A0701	Capturing Problem	26/11/2021	26/11/2021
									A070103	Intermittent Capture	26/11/2021	26/11/2021
									A1502	Positioning Problem	26/11/2021	26/11/2021
							E0602	Cardiac Arrest	A0701	Capturing Problem	26/11/2021	26/11/2021
									A070103	Intermittent Capture	26/11/2021	26/11/2021
									A1502	Positioning Problem	26/11/2021	26/11/2021
							E2403	No Clinical Signs, Symptoms or Conditions	A0701	Capturing Problem	26/11/2021	26/11/2021
									A070103	Intermittent Capture	26/11/2021	26/11/2021
									A1502	Positioning Problem	26/11/2021	26/11/2021
							F08	Hospitalization or Prolonged Hospitalization	A0701	Capturing Problem	26/11/2021	26/11/2021
									A070103	Intermittent Capture	26/11/2021	26/11/2021
									A1502	Positioning Problem	26/11/2021	26/11/2021
							F1203	Life Threatening Illness or Injury	A0701	Capturing Problem	26/11/2021	26/11/2021
									A070103	Intermittent Capture	26/11/2021	26/11/2021
									A1502	Positioning Problem	26/11/2021	26/11/2021
							F23	Unexpected Medical Intervention	A0701	Capturing Problem	26/11/2021	26/11/2021

									A070103	Intermittent Capture	26/11/2021	26/11/2021
									A1502	Positioning Problem	26/11/2021	26/11/2021
7215	2021	All other reportable incidents	17882	Physiologic-monitoring defibrillation system	CT1684	Cardiac pacemakers and associated devices	E2401	Insufficient Information	A090801	Signal Artifact/Noise		
							E2403	No Clinical Signs, Symptoms or Conditions	A090801	Signal Artifact/Noise		
							F26	No Health Consequences or Impact	A090801	Signal Artifact/Noise		
7216	2021	All other reportable incidents	36241	Pacemaker/defibrillator lead	CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A0203	Defective Device		
									A070909	Oversensing		
							F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A0203	Defective Device		
									A070909	Oversensing		
7217	2021	All other reportable incidents	36241	Pacemaker/defibrillator lead	CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A0203	Defective Device		
									A070909	Oversensing		
							F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A0203	Defective Device		
									A070909	Oversensing		
7218	2021	All other reportable incidents	47265	Dual-chamber implantable pacemaker, rate-responsive	CT141	Electrodes/Leads and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem	14/02/2022	14/02/2022
									A070504	Premature Discharge of Battery	14/02/2022	14/02/2022
							F08	Hospitalization or Prolonged Hospitalization	A0705	Battery Problem	14/02/2022	14/02/2022
									A070504	Premature Discharge of Battery	14/02/2022	14/02/2022
							F1905	Device Revision or Replacement	A0705	Battery Problem	14/02/2022	14/02/2022
									A070504	Premature Discharge of Battery	14/02/2022	14/02/2022
					CT145	Cardiovascular devices	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem	14/02/2022	14/02/2022
									A070504	Premature Discharge of Battery	14/02/2022	14/02/2022
							F08	Hospitalization or Prolonged Hospitalization	A0705	Battery Problem	14/02/2022	14/02/2022
									A070504	Premature Discharge of Battery	14/02/2022	14/02/2022
							F1905	Device Revision or Replacement	A0705	Battery Problem	14/02/2022	14/02/2022
									A070504	Premature Discharge of Battery	14/02/2022	14/02/2022
					CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem	14/02/2022	14/02/2022
									A070504	Premature Discharge of Battery	14/02/2022	14/02/2022
							F08	Hospitalization or Prolonged Hospitalization	A0705	Battery Problem	14/02/2022	14/02/2022
									A070504	Premature Discharge of Battery	14/02/2022	14/02/2022
							F1905	Device Revision or Replacement	A0705	Battery Problem	14/02/2022	14/02/2022
									A070504	Premature Discharge of Battery	14/02/2022	14/02/2022
					CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem	14/02/2022	14/02/2022
									A070504	Premature Discharge of Battery	14/02/2022	14/02/2022
							F08	Hospitalization or Prolonged Hospitalization	A0705	Battery Problem	14/02/2022	14/02/2022
									A070504	Premature Discharge of Battery	14/02/2022	14/02/2022
							F1905	Device Revision or Replacement	A0705	Battery Problem	14/02/2022	14/02/2022
									A070504	Premature Discharge of Battery	14/02/2022	14/02/2022
					CT233	Surgical	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem	14/02/2022	14/02/2022
									A070504	Premature Discharge of Battery	14/02/2022	14/02/2022
							F08	Hospitalization or Prolonged Hospitalization	A0705	Battery Problem	14/02/2022	14/02/2022
									A070504	Premature Discharge of Battery	14/02/2022	14/02/2022
							F1905	Device Revision or Replacement	A0705	Battery Problem	14/02/2022	14/02/2022
									A070504	Premature Discharge of Battery	14/02/2022	14/02/2022
					CT2406	Implantable	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem	14/02/2022	14/02/2022
									A070504	Premature Discharge of Battery	14/02/2022	14/02/2022
							F08	Hospitalization or Prolonged Hospitalization	A0705	Battery Problem	14/02/2022	14/02/2022
									A070504	Premature Discharge of Battery	14/02/2022	14/02/2022
							F1905	Device Revision or Replacement	A0705	Battery Problem	14/02/2022	14/02/2022
									A070504	Premature Discharge of Battery	14/02/2022	14/02/2022
					CT2407	Active implantable	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem	14/02/2022	14/02/2022
									A070504	Premature Discharge of Battery	14/02/2022	14/02/2022
							F08	Hospitalization or Prolonged Hospitalization	A0705	Battery Problem	14/02/2022	14/02/2022
									A070504	Premature Discharge of Battery	14/02/2022	14/02/2022
							F1905	Device Revision or Replacement	A0705	Battery Problem	14/02/2022	14/02/2022
									A070504	Premature Discharge of Battery	14/02/2022	14/02/2022
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem	14/02/2022	14/02/2022
									A070504	Premature Discharge of Battery	14/02/2022	14/02/2022
							F08	Hospitalization or Prolonged Hospitalization	A0705	Battery Problem	14/02/2022	14/02/2022
									A070504	Premature Discharge of Battery	14/02/2022	14/02/2022
							F1905	Device Revision or Replacement	A0705	Battery Problem	14/02/2022	14/02/2022
									A070504	Premature Discharge of Battery	14/02/2022	14/02/2022
					CT2727	Active implant devices/ancillaries	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem	14/02/2022	14/02/2022
									A070504	Premature Discharge of Battery	14/02/2022	14/02/2022
							F08	Hospitalization or Prolonged Hospitalization	A0705	Battery Problem	14/02/2022	14/02/2022
									A070504	Premature Discharge of Battery	14/02/2022	14/02/2022
							F1905	Device Revision or Replacement	A0705	Battery Problem	14/02/2022	14/02/2022
									A070504	Premature Discharge of Battery	14/02/2022	14/02/2022
					CT2750	Implantable cardiac electrical stimulators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem	14/02/2022	14/02/2022
									A070504	Premature Discharge of Battery	14/02/2022	14/02/2022
							F08	Hospitalization or Prolonged Hospitalization	A0705	Battery Problem	14/02/2022	14/02/2022
									A070504	Premature Discharge of Battery	14/02/2022	14/02/2022
							F1905	Device Revision or Replacement	A0705	Battery Problem	14/02/2022	14/02/2022
									A070504	Premature Discharge of Battery	14/02/2022	14/02/2022
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem	14/02/2022	14/02/2022
									A070504	Premature Discharge of Battery	14/02/2022	14/02/2022
							F08	Hospitalization or Prolonged Hospitalization	A0705	Battery Problem	14/02/2022	14/02/2022
									A070504	Premature Discharge of Battery	14/02/2022	14/02/2022
							F1905	Device Revision or Replacement	A0705	Battery Problem	14/02/2022	14/02/2022
									A070504	Premature Discharge of Battery	14/02/2022	14/02/2022
					CT2985	Active	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem	14/02/2022	14/02/2022
									A070504	Premature Discharge of Battery	14/02/2022	14/02/2022
							F08	Hospitalization or Prolonged Hospitalization	A0705	Battery Problem	14/02/2022	14/02/2022
									A070504	Premature Discharge of Battery	14/02/2022	14/02/2022
							F1905	Device Revision or Replacement	A0705	Battery Problem	14/02/2022	14/02/2022
									A070504	Premature Discharge of Battery	14/02/2022	14/02/2022
					CT3001	Implantable metal	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem	14/02/2022	14/02/2022
									A070504	Premature Discharge of Battery	14/02/2022	14/02/2022
							F08	Hospitalization or Prolonged Hospitalization	A0705	Battery Problem	14/02/2022	14/02/2022

									A070504	Premature Discharge of Battery	14/02/2022	14/02/2022
						F1905	Device Revision or Replacement	A0705	Battery Problem	14/02/2022	14/02/2022
								A070504	Premature Discharge of Battery	14/02/2022	14/02/2022
					CT3002	Implantable synthetic polymer	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem	14/02/2022	14/02/2022
									A070504	Premature Discharge of Battery	14/02/2022	14/02/2022
							F08	Hospitalization or Prolonged Hospitalization	A0705	Battery Problem	14/02/2022	14/02/2022
									A070504	Premature Discharge of Battery	14/02/2022	14/02/2022
							F1905	Device Revision or Replacement	A0705	Battery Problem	14/02/2022	14/02/2022
									A070504	Premature Discharge of Battery	14/02/2022	14/02/2022
					CT3006	Implantable inorganic material	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem	14/02/2022	14/02/2022
									A070504	Premature Discharge of Battery	14/02/2022	14/02/2022
							F08	Hospitalization or Prolonged Hospitalization	A0705	Battery Problem	14/02/2022	14/02/2022
									A070504	Premature Discharge of Battery	14/02/2022	14/02/2022
							F1905	Device Revision or Replacement	A0705	Battery Problem	14/02/2022	14/02/2022
									A070504	Premature Discharge of Battery	14/02/2022	14/02/2022
					CT314	Electrically powered	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem	14/02/2022	14/02/2022
									A070504	Premature Discharge of Battery	14/02/2022	14/02/2022
							F08	Hospitalization or Prolonged Hospitalization	A0705	Battery Problem	14/02/2022	14/02/2022
									A070504	Premature Discharge of Battery	14/02/2022	14/02/2022
							F1905	Device Revision or Replacement	A0705	Battery Problem	14/02/2022	14/02/2022
									A070504	Premature Discharge of Battery	14/02/2022	14/02/2022
					CT321	Long-term surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem	14/02/2022	14/02/2022
									A070504	Premature Discharge of Battery	14/02/2022	14/02/2022
							F08	Hospitalization or Prolonged Hospitalization	A0705	Battery Problem	14/02/2022	14/02/2022
									A070504	Premature Discharge of Battery	14/02/2022	14/02/2022
							F1905	Device Revision or Replacement	A0705	Battery Problem	14/02/2022	14/02/2022
									A070504	Premature Discharge of Battery	14/02/2022	14/02/2022
					CT449	Generators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem	14/02/2022	14/02/2022
									A070504	Premature Discharge of Battery	14/02/2022	14/02/2022
							F08	Hospitalization or Prolonged Hospitalization	A0705	Battery Problem	14/02/2022	14/02/2022
									A070504	Premature Discharge of Battery	14/02/2022	14/02/2022
							F1905	Device Revision or Replacement	A0705	Battery Problem	14/02/2022	14/02/2022
									A070504	Premature Discharge of Battery	14/02/2022	14/02/2022
					CT981	Single use	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem	14/02/2022	14/02/2022
									A070504	Premature Discharge of Battery	14/02/2022	14/02/2022
							F08	Hospitalization or Prolonged Hospitalization	A0705	Battery Problem	14/02/2022	14/02/2022
									A070504	Premature Discharge of Battery	14/02/2022	14/02/2022
							F1905	Device Revision or Replacement	A0705	Battery Problem	14/02/2022	14/02/2022
									A070504	Premature Discharge of Battery	14/02/2022	14/02/2022
					CT983	Surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem	14/02/2022	14/02/2022
									A070504	Premature Discharge of Battery	14/02/2022	14/02/2022
							F08	Hospitalization or Prolonged Hospitalization	A0705	Battery Problem	14/02/2022	14/02/2022
									A070504	Premature Discharge of Battery	14/02/2022	14/02/2022
							F1905	Device Revision or Replacement	A0705	Battery Problem	14/02/2022	14/02/2022
									A070504	Premature Discharge of Battery	14/02/2022	14/02/2022
7219	2021	All other reportable incidents	35852	Single-chamber implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem	13/01/2022	13/01/2022
									A070504	Premature Discharge of Battery	13/01/2022	13/01/2022
							F08	Hospitalization or Prolonged Hospitalization	A0705	Battery Problem	13/01/2022	13/01/2022
									A070504	Premature Discharge of Battery	13/01/2022	13/01/2022
							F1903	Device Explantation	A0705	Battery Problem	13/01/2022	13/01/2022
									A070504	Premature Discharge of Battery	13/01/2022	13/01/2022
7220	2021	All other reportable incidents	47263	Cardiac resynchronization therapy implantable pacemaker	CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A0908	Incorrect, Inadequate or Imprecise Result or Readings	07/10/2021	07/10/2021
							F08	Hospitalization or Prolonged Hospitalization	A0908	Incorrect, Inadequate or Imprecise Result or Readings	07/10/2021	07/10/2021
							F1905	Device Revision or Replacement	A0908	Incorrect, Inadequate or Imprecise Result or Readings	07/10/2021	07/10/2021
					CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A0908	Incorrect, Inadequate or Imprecise Result or Readings	07/10/2021	07/10/2021
							F08	Hospitalization or Prolonged Hospitalization	A0908	Incorrect, Inadequate or Imprecise Result or Readings	07/10/2021	07/10/2021
							F1905	Device Revision or Replacement	A0908	Incorrect, Inadequate or Imprecise Result or Readings	07/10/2021	07/10/2021
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A0908	Incorrect, Inadequate or Imprecise Result or Readings	07/10/2021	07/10/2021
							F08	Hospitalization or Prolonged Hospitalization	A0908	Incorrect, Inadequate or Imprecise Result or Readings	07/10/2021	07/10/2021
							F1905	Device Revision or Replacement	A0908	Incorrect, Inadequate or Imprecise Result or Readings	07/10/2021	07/10/2021
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A0908	Incorrect, Inadequate or Imprecise Result or Readings	07/10/2021	07/10/2021
							F08	Hospitalization or Prolonged Hospitalization	A0908	Incorrect, Inadequate or Imprecise Result or Readings	07/10/2021	07/10/2021
							F1905	Device Revision or Replacement	A0908	Incorrect, Inadequate or Imprecise Result or Readings	07/10/2021	07/10/2021
7221	2022	All other reportable incidents	36241	Pacemaker/defibrillator lead	CT1684	Cardiac pacemakers and associated devices	E2104	Electric Shock	A070909	Oversensing		
									A0712	Pacing Problem		
									A071302	Inappropriate/Inadequate Shock/Stimulation		
									A090801	Signal Artifact/Noise		
							E2403	No Clinical Signs, Symptoms or Conditions	A070909	Oversensing		
									A0712	Pacing Problem		
									A071302	Inappropriate/Inadequate Shock/Stimulation		
									A090801	Signal Artifact/Noise		
							F08	Hospitalization or Prolonged Hospitalization	A070909	Oversensing		
									A0712	Pacing Problem		
									A071302	Inappropriate/Inadequate Shock/Stimulation		
									A090801	Signal Artifact/Noise		
							F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A070909	Oversensing		
									A0712	Pacing Problem		
									A071302	Inappropriate/Inadequate Shock/Stimulation		
									A090801	Signal Artifact/Noise		

							F1905	Device Revision or Replacement	A070909	Oversensing			
									A0712	Pacing Problem			
									A071302	Inappropriate/Inadequate Shock/Stimulation			
									A090801	Signal Artifact/Noise			
7222	2022	All other reportable incidents	47263	Cardiac resynchronization therapy implantable pacemaker	CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A0908	Incorrect, Inadequate or Imprecise Result or Readings	18/07/2022	18/07/2022
							F08	Hospitalization or Prolonged Hospitalization	A0908	Incorrect, Inadequate or Imprecise Result or Readings	18/07/2022	18/07/2022
							F1905	Device Revision or Replacement	A0908	Incorrect, Inadequate or Imprecise Result or Readings	18/07/2022	18/07/2022
					CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A0908	Incorrect, Inadequate or Imprecise Result or Readings	18/07/2022	18/07/2022
							F08	Hospitalization or Prolonged Hospitalization	A0908	Incorrect, Inadequate or Imprecise Result or Readings	18/07/2022	18/07/2022
							F1905	Device Revision or Replacement	A0908	Incorrect, Inadequate or Imprecise Result or Readings	18/07/2022	18/07/2022
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A0908	Incorrect, Inadequate or Imprecise Result or Readings	18/07/2022	18/07/2022
							F08	Hospitalization or Prolonged Hospitalization	A0908	Incorrect, Inadequate or Imprecise Result or Readings	18/07/2022	18/07/2022
							F1905	Device Revision or Replacement	A0908	Incorrect, Inadequate or Imprecise Result or Readings	18/07/2022	18/07/2022
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A0908	Incorrect, Inadequate or Imprecise Result or Readings	18/07/2022	18/07/2022
							F08	Hospitalization or Prolonged Hospitalization	A0908	Incorrect, Inadequate or Imprecise Result or Readings	18/07/2022	18/07/2022
							F1905	Device Revision or Replacement	A0908	Incorrect, Inadequate or Imprecise Result or Readings	18/07/2022	18/07/2022
7223	2022	All other reportable incidents	47265	Dual-chamber implantable pacemaker, rate-responsive	CT141	Electrodes/Leads and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	23/08/2022	23/08/2022
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	23/08/2022	23/08/2022
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	23/08/2022	23/08/2022
					CT145	Cardiovascular devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	23/08/2022	23/08/2022
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	23/08/2022	23/08/2022
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	23/08/2022	23/08/2022
					CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	23/08/2022	23/08/2022
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	23/08/2022	23/08/2022
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	23/08/2022	23/08/2022
					CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	23/08/2022	23/08/2022
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	23/08/2022	23/08/2022
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	23/08/2022	23/08/2022
					CT233	Surgical	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	23/08/2022	23/08/2022
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	23/08/2022	23/08/2022
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	23/08/2022	23/08/2022
					CT2406	Implantable	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	23/08/2022	23/08/2022
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	23/08/2022	23/08/2022
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	23/08/2022	23/08/2022
					CT2407	Active implantable	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	23/08/2022	23/08/2022
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	23/08/2022	23/08/2022
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	23/08/2022	23/08/2022
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	23/08/2022	23/08/2022
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	23/08/2022	23/08/2022
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	23/08/2022	23/08/2022
					CT2727	Active implant devices/ancillaries	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	23/08/2022	23/08/2022
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	23/08/2022	23/08/2022
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	23/08/2022	23/08/2022
					CT2750	Implantable cardiac electrical stimulators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	23/08/2022	23/08/2022
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	23/08/2022	23/08/2022
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	23/08/2022	23/08/2022
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	23/08/2022	23/08/2022
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	23/08/2022	23/08/2022
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	23/08/2022	23/08/2022
					CT2985	Active	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	23/08/2022	23/08/2022
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	23/08/2022	23/08/2022
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	23/08/2022	23/08/2022
					CT3001	Implantable metal	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	23/08/2022	23/08/2022
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	23/08/2022	23/08/2022
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	23/08/2022	23/08/2022
					CT3002	Implantable synthetic polymer	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	23/08/2022	23/08/2022
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	23/08/2022	23/08/2022
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	23/08/2022	23/08/2022
					CT3006	Implantable inorganic material	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	23/08/2022	23/08/2022
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	23/08/2022	23/08/2022
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	23/08/2022	23/08/2022
					CT314	Electrically powered	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	23/08/2022	23/08/2022
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	23/08/2022	23/08/2022
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	23/08/2022	23/08/2022
					CT321	Long-term surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	23/08/2022	23/08/2022
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	23/08/2022	23/08/2022
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	23/08/2022	23/08/2022
					CT449	Generators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	23/08/2022	23/08/2022
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	23/08/2022	23/08/2022
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	23/08/2022	23/08/2022
					CT981	Single use	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	23/08/2022	23/08/2022
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	23/08/2022	23/08/2022
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	23/08/2022	23/08/2022
					CT983	Surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	23/08/2022	23/08/2022
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	23/08/2022	23/08/2022
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	23/08/2022	23/08/2022
7224	2022	All other reportable incidents	35223	Endocardial/interventricular septal pacing lead	CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070901	Decreased Sensitivity		
									A070909	Oversensing		
									A090801	Signal Artifact/Noise		

							F26	No Health Consequences or Impact	A070901	Decreased Sensitivity		
									A070909	Oversensing		
									A090801	Signal Artifact/Noise		
7225	2022	All other reportable incidents	17882	Physiologic-monitoring defibrillation system	CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070803	Failure to Power Up		
							F27	Problem Identified During Non-Clinical Procedure	A070803	Failure to Power Up		
7226	2022	All other reportable incidents	35223	Endocardial/interventricular septal pacing lead	CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070910	Undersensing		
							F26	No Health Consequences or Impact	A070910	Undersensing		
7227	2021	All other reportable incidents	47270	Cardiac resynchronization therapy implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E1906	Unspecified Infection	A070909	Oversensing	26/11/2021	26/11/2021
									A0712	Pacing Problem	26/11/2021	26/11/2021
									A071204	No Pacing	26/11/2021	26/11/2021
							E2403	No Clinical Signs, Symptoms or Conditions	A070909	Oversensing	26/11/2021	26/11/2021
									A0712	Pacing Problem	26/11/2021	26/11/2021
									A071204	No Pacing	26/11/2021	26/11/2021
							F26	No Health Consequences or Impact	A070909	Oversensing	26/11/2021	26/11/2021
									A0712	Pacing Problem	26/11/2021	26/11/2021
									A071204	No Pacing	26/11/2021	26/11/2021
7228	2022	All other reportable incidents	37265	Dual-chamber implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem	13/07/2022	13/07/2022
									A070504	Premature Discharge of Battery	13/07/2022	13/07/2022
									A1304	Telemetry Discrepancy	13/07/2022	13/07/2022
							F08	Hospitalization or Prolonged Hospitalization	A0705	Battery Problem	13/07/2022	13/07/2022
									A070504	Premature Discharge of Battery	13/07/2022	13/07/2022
									A1304	Telemetry Discrepancy	13/07/2022	13/07/2022
							F1905	Device Revision or Replacement	A0705	Battery Problem	13/07/2022	13/07/2022
									A070504	Premature Discharge of Battery	13/07/2022	13/07/2022
									A1304	Telemetry Discrepancy	13/07/2022	13/07/2022
7229	2021	All other reportable incidents	35852	Single-chamber implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E2104	Electric Shock	A070909	Oversensing		
									A071302	Inappropriate/Inadequate Shock/Stimulation		
							E2403	No Clinical Signs, Symptoms or Conditions	A070909	Oversensing		
									A071302	Inappropriate/Inadequate Shock/Stimulation		
							F26	No Health Consequences or Impact	A070909	Oversensing		
									A071302	Inappropriate/Inadequate Shock/Stimulation		
7230	2021	All other reportable incidents	36241	Pacemaker/defibrillator lead	CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070102	High Capture Threshold		
									A072201	High Impedance		
							F08	Hospitalization or Prolonged Hospitalization	A070102	High Capture Threshold		
									A072201	High Impedance		
							F17	Sedation	A070102	High Capture Threshold		
									A072201	High Impedance		
							F22	Unexpected Diagnostic Intervention	A070102	High Capture Threshold		
									A072201	High Impedance		
7231	2021	All other reportable incidents	36241	Pacemaker/defibrillator lead	CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A0203	Defective Device	15/11/2021	15/11/2021
									A072201	High Impedance	15/11/2021	15/11/2021
									A090801	Signal Artifact/Noise	15/11/2021	15/11/2021
							F08	Hospitalization or Prolonged Hospitalization	A0203	Defective Device	15/11/2021	15/11/2021
									A072201	High Impedance	15/11/2021	15/11/2021
									A090801	Signal Artifact/Noise	15/11/2021	15/11/2021
							F1905	Device Revision or Replacement	A0203	Defective Device	15/11/2021	15/11/2021
									A072201	High Impedance	15/11/2021	15/11/2021
									A090801	Signal Artifact/Noise	15/11/2021	15/11/2021
7232	2021	All other reportable incidents	47265	Dual-chamber implantable pacemaker, rate-responsive	CT141	Electrodes/Leads and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem	20/12/2021	20/12/2021
									A070504	Premature Discharge of Battery	20/12/2021	20/12/2021
							F08	Hospitalization or Prolonged Hospitalization	A0705	Battery Problem	20/12/2021	20/12/2021
									A070504	Premature Discharge of Battery	20/12/2021	20/12/2021
							F1905	Device Revision or Replacement	A0705	Battery Problem	20/12/2021	20/12/2021
									A070504	Premature Discharge of Battery	20/12/2021	20/12/2021
					CT145	Cardiovascular devices	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem	20/12/2021	20/12/2021
									A070504	Premature Discharge of Battery	20/12/2021	20/12/2021
							F08	Hospitalization or Prolonged Hospitalization	A0705	Battery Problem	20/12/2021	20/12/2021
									A070504	Premature Discharge of Battery	20/12/2021	20/12/2021
							F1905	Device Revision or Replacement	A0705	Battery Problem	20/12/2021	20/12/2021
									A070504	Premature Discharge of Battery	20/12/2021	20/12/2021
					CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem	20/12/2021	20/12/2021
									A070504	Premature Discharge of Battery	20/12/2021	20/12/2021
							F08	Hospitalization or Prolonged Hospitalization	A0705	Battery Problem	20/12/2021	20/12/2021
									A070504	Premature Discharge of Battery	20/12/2021	20/12/2021
							F1905	Device Revision or Replacement	A0705	Battery Problem	20/12/2021	20/12/2021
									A070504	Premature Discharge of Battery	20/12/2021	20/12/2021
					CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem	20/12/2021	20/12/2021
									A070504	Premature Discharge of Battery	20/12/2021	20/12/2021
							F08	Hospitalization or Prolonged Hospitalization	A0705	Battery Problem	20/12/2021	20/12/2021
									A070504	Premature Discharge of Battery	20/12/2021	20/12/2021
							F1905	Device Revision or Replacement	A0705	Battery Problem	20/12/2021	20/12/2021
									A070504	Premature Discharge of Battery	20/12/2021	20/12/2021
					CT233	Surgical	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem	20/12/2021	20/12/2021
									A070504	Premature Discharge of Battery	20/12/2021	20/12/2021
							F08	Hospitalization or Prolonged Hospitalization	A0705	Battery Problem	20/12/2021	20/12/2021
									A070504	Premature Discharge of Battery	20/12/2021	20/12/2021
							F1905	Device Revision or Replacement	A0705	Battery Problem	20/12/2021	20/12/2021
									A070504	Premature Discharge of Battery	20/12/2021	20/12/2021
					CT2406	Implantable	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem	20/12/2021	20/12/2021
									A070504	Premature Discharge of Battery	20/12/2021	20/12/2021
							F08	Hospitalization or Prolonged Hospitalization	A0705	Battery Problem	20/12/2021	20/12/2021
									A070504	Premature Discharge of Battery	20/12/2021	20/12/2021
							F1905	Device Revision or Replacement	A0705	Battery Problem	20/12/2021	20/12/2021
									A070504	Premature Discharge of Battery	20/12/2021	20/12/2021
					CT2407	Active implantable	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem	20/12/2021	20/12/2021
									A070504	Premature Discharge of Battery	20/12/2021	20/12/2021
							F08	Hospitalization or Prolonged Hospitalization	A0705	Battery Problem	20/12/2021	20/12/2021
									A070504	Premature Discharge of Battery	20/12/2021	20/12/2021

							F1905	Device Revision or Replacement	A0705	Battery Problem	20/12/2021	20/12/2021
									A070504	Premature Discharge of Battery	20/12/2021	20/12/2021
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem	20/12/2021	20/12/2021
									A070504	Premature Discharge of Battery	20/12/2021	20/12/2021
							F08	Hospitalization or Prolonged Hospitalization	A0705	Battery Problem	20/12/2021	20/12/2021
									A070504	Premature Discharge of Battery	20/12/2021	20/12/2021
							F1905	Device Revision or Replacement	A0705	Battery Problem	20/12/2021	20/12/2021
									A070504	Premature Discharge of Battery	20/12/2021	20/12/2021
					CT2727	Active implant devices/ancillaries	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem	20/12/2021	20/12/2021
									A070504	Premature Discharge of Battery	20/12/2021	20/12/2021
							F08	Hospitalization or Prolonged Hospitalization	A0705	Battery Problem	20/12/2021	20/12/2021
									A070504	Premature Discharge of Battery	20/12/2021	20/12/2021
							F1905	Device Revision or Replacement	A0705	Battery Problem	20/12/2021	20/12/2021
									A070504	Premature Discharge of Battery	20/12/2021	20/12/2021
					CT2750	Implantable cardiac electrical stimulators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem	20/12/2021	20/12/2021
									A070504	Premature Discharge of Battery	20/12/2021	20/12/2021
							F08	Hospitalization or Prolonged Hospitalization	A0705	Battery Problem	20/12/2021	20/12/2021
									A070504	Premature Discharge of Battery	20/12/2021	20/12/2021
							F1905	Device Revision or Replacement	A0705	Battery Problem	20/12/2021	20/12/2021
									A070504	Premature Discharge of Battery	20/12/2021	20/12/2021
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem	20/12/2021	20/12/2021
									A070504	Premature Discharge of Battery	20/12/2021	20/12/2021
							F08	Hospitalization or Prolonged Hospitalization	A0705	Battery Problem	20/12/2021	20/12/2021
									A070504	Premature Discharge of Battery	20/12/2021	20/12/2021
							F1905	Device Revision or Replacement	A0705	Battery Problem	20/12/2021	20/12/2021
									A070504	Premature Discharge of Battery	20/12/2021	20/12/2021
					CT2985	Active	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem	20/12/2021	20/12/2021
									A070504	Premature Discharge of Battery	20/12/2021	20/12/2021
							F08	Hospitalization or Prolonged Hospitalization	A0705	Battery Problem	20/12/2021	20/12/2021
									A070504	Premature Discharge of Battery	20/12/2021	20/12/2021
							F1905	Device Revision or Replacement	A0705	Battery Problem	20/12/2021	20/12/2021
									A070504	Premature Discharge of Battery	20/12/2021	20/12/2021
					CT3001	Implantable metal	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem	20/12/2021	20/12/2021
									A070504	Premature Discharge of Battery	20/12/2021	20/12/2021
							F08	Hospitalization or Prolonged Hospitalization	A0705	Battery Problem	20/12/2021	20/12/2021
									A070504	Premature Discharge of Battery	20/12/2021	20/12/2021
							F1905	Device Revision or Replacement	A0705	Battery Problem	20/12/2021	20/12/2021
									A070504	Premature Discharge of Battery	20/12/2021	20/12/2021
					CT3002	Implantable synthetic polymer	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem	20/12/2021	20/12/2021
									A070504	Premature Discharge of Battery	20/12/2021	20/12/2021
							F08	Hospitalization or Prolonged Hospitalization	A0705	Battery Problem	20/12/2021	20/12/2021
									A070504	Premature Discharge of Battery	20/12/2021	20/12/2021
							F1905	Device Revision or Replacement	A0705	Battery Problem	20/12/2021	20/12/2021
									A070504	Premature Discharge of Battery	20/12/2021	20/12/2021
					CT3006	Implantable inorganic material	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem	20/12/2021	20/12/2021
									A070504	Premature Discharge of Battery	20/12/2021	20/12/2021
							F08	Hospitalization or Prolonged Hospitalization	A0705	Battery Problem	20/12/2021	20/12/2021
									A070504	Premature Discharge of Battery	20/12/2021	20/12/2021
							F1905	Device Revision or Replacement	A0705	Battery Problem	20/12/2021	20/12/2021
									A070504	Premature Discharge of Battery	20/12/2021	20/12/2021
					CT314	Electrically powered	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem	20/12/2021	20/12/2021
									A070504	Premature Discharge of Battery	20/12/2021	20/12/2021
							F08	Hospitalization or Prolonged Hospitalization	A0705	Battery Problem	20/12/2021	20/12/2021
									A070504	Premature Discharge of Battery	20/12/2021	20/12/2021
							F1905	Device Revision or Replacement	A0705	Battery Problem	20/12/2021	20/12/2021
									A070504	Premature Discharge of Battery	20/12/2021	20/12/2021
					CT321	Long-term surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem	20/12/2021	20/12/2021
									A070504	Premature Discharge of Battery	20/12/2021	20/12/2021
							F08	Hospitalization or Prolonged Hospitalization	A0705	Battery Problem	20/12/2021	20/12/2021
									A070504	Premature Discharge of Battery	20/12/2021	20/12/2021
							F1905	Device Revision or Replacement	A0705	Battery Problem	20/12/2021	20/12/2021
									A070504	Premature Discharge of Battery	20/12/2021	20/12/2021
					CT449	Generators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem	20/12/2021	20/12/2021
									A070504	Premature Discharge of Battery	20/12/2021	20/12/2021
							F08	Hospitalization or Prolonged Hospitalization	A0705	Battery Problem	20/12/2021	20/12/2021
									A070504	Premature Discharge of Battery	20/12/2021	20/12/2021
							F1905	Device Revision or Replacement	A0705	Battery Problem	20/12/2021	20/12/2021
									A070504	Premature Discharge of Battery	20/12/2021	20/12/2021
					CT981	Single use	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem	20/12/2021	20/12/2021
									A070504	Premature Discharge of Battery	20/12/2021	20/12/2021
							F08	Hospitalization or Prolonged Hospitalization	A0705	Battery Problem	20/12/2021	20/12/2021
									A070504	Premature Discharge of Battery	20/12/2021	20/12/2021
							F1905	Device Revision or Replacement	A0705	Battery Problem	20/12/2021	20/12/2021
									A070504	Premature Discharge of Battery	20/12/2021	20/12/2021
					CT983	Surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem	20/12/2021	20/12/2021
									A070504	Premature Discharge of Battery	20/12/2021	20/12/2021
							F08	Hospitalization or Prolonged Hospitalization	A0705	Battery Problem	20/12/2021	20/12/2021
									A070504	Premature Discharge of Battery	20/12/2021	20/12/2021
							F1905	Device Revision or Replacement	A0705	Battery Problem	20/12/2021	20/12/2021
									A070504	Premature Discharge of Battery	20/12/2021	20/12/2021
7233	2022	All other reportable incidents	47205	Cardiac pulse generator programmer	CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A090805	Incorrect Measurement		
							F23	Unexpected Medical Intervention	A090805	Incorrect Measurement		
7234	2021	All other reportable incidents	35223	Endocardial/interventricular septal pacing lead	CT1684	Cardiac pacemakers and associated devices	E011903	Syncope/Fainting	A0401	Break		
									A070101	Failure to Capture		
									A072202	Low Impedance		
							E2403	No Clinical Signs, Symptoms or Conditions	A0401	Break		
									A070101	Failure to Capture		
									A072202	Low Impedance		

							F12	Serious Injury/ Illness/ Impairment	A0401	Break		
									A070101	Failure to Capture		
									A072202	Low Impedance		
7235	2022	All other reportable incidents	35223	Endocardial/interventricular septal pacing lead	CT1684	Cardiac pacemakers and associated devices	E0602	Cardiac Arrest	A0712	Pacing Problem		
							F1203	Life Threatening Illness or Injury	A0712	Pacing Problem		
7236	2022	All other reportable incidents	47265	Dual-chamber implantable pacemaker, rate-responsive	CT141	Electrodes/Leads and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	21/03/2022	21/03/2022
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	21/03/2022	21/03/2022
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	21/03/2022	21/03/2022
					CT145	Cardiovascular devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	21/03/2022	21/03/2022
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	21/03/2022	21/03/2022
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	21/03/2022	21/03/2022
					CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	21/03/2022	21/03/2022
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	21/03/2022	21/03/2022
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	21/03/2022	21/03/2022
					CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	21/03/2022	21/03/2022
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	21/03/2022	21/03/2022
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	21/03/2022	21/03/2022
					CT233	Surgical	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	21/03/2022	21/03/2022
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	21/03/2022	21/03/2022
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	21/03/2022	21/03/2022
					CT2406	Implantable	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	21/03/2022	21/03/2022
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	21/03/2022	21/03/2022
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	21/03/2022	21/03/2022
					CT2407	Active implantable	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	21/03/2022	21/03/2022
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	21/03/2022	21/03/2022
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	21/03/2022	21/03/2022
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	21/03/2022	21/03/2022
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	21/03/2022	21/03/2022
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	21/03/2022	21/03/2022
					CT2727	Active implant devices/ancillaries	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	21/03/2022	21/03/2022
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	21/03/2022	21/03/2022
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	21/03/2022	21/03/2022
					CT2750	Implantable cardiac electrical stimulators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	21/03/2022	21/03/2022
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	21/03/2022	21/03/2022
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	21/03/2022	21/03/2022
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	21/03/2022	21/03/2022
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	21/03/2022	21/03/2022
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	21/03/2022	21/03/2022
					CT2985	Active	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	21/03/2022	21/03/2022
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	21/03/2022	21/03/2022
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	21/03/2022	21/03/2022
					CT3001	Implantable metal	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	21/03/2022	21/03/2022
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	21/03/2022	21/03/2022
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	21/03/2022	21/03/2022
					CT3002	Implantable synthetic polymer	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	21/03/2022	21/03/2022
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	21/03/2022	21/03/2022
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	21/03/2022	21/03/2022
					CT3006	Implantable inorganic material	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	21/03/2022	21/03/2022
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	21/03/2022	21/03/2022
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	21/03/2022	21/03/2022
					CT314	Electrically powered	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	21/03/2022	21/03/2022
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	21/03/2022	21/03/2022
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	21/03/2022	21/03/2022
					CT321	Long-term surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	21/03/2022	21/03/2022
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	21/03/2022	21/03/2022
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	21/03/2022	21/03/2022
					CT449	Generators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	21/03/2022	21/03/2022
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	21/03/2022	21/03/2022
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	21/03/2022	21/03/2022
					CT981	Single use	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	21/03/2022	21/03/2022
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	21/03/2022	21/03/2022
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	21/03/2022	21/03/2022
					CT983	Surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	21/03/2022	21/03/2022
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	21/03/2022	21/03/2022
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	21/03/2022	21/03/2022
7237	2021	All other reportable incidents	17882	Physiologic-monitoring defibrillation system	CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070803	Failure to Power Up		
							F27	Problem Identified During Non-Clinical Procedure	A070803	Failure to Power Up		
7238	2022	All other reportable incidents	35852	Single-chamber implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	22/06/2022	22/06/2022
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	22/06/2022	22/06/2022
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	22/06/2022	22/06/2022
7239	2021	All other reportable incidents	17882	Physiologic-monitoring defibrillation system	CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A07	Electrical /Electronic Property Problem		
							F27	Problem Identified During Non-Clinical Procedure	A07	Electrical /Electronic Property Problem		
7240	2021	All other reportable incidents	46222	Pacing/defibrillation lead cap kit	CT1684	Cardiac pacemakers and associated devices	E200801	Device Embedded In Tissue or Plaque	A0501	Detachment of Device or Device Component		
							E2403	No Clinical Signs, Symptoms or Conditions	A0501	Detachment of Device or Device Component		
							F28	Appropriate Health Impact Term/Code not Available	A0501	Detachment of Device or Device Component		
7241	2021	All other reportable incidents	37265	Dual-chamber implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem	02/12/2021	02/12/2021
									A1107	Data Problem	02/12/2021	02/12/2021
							F08	Hospitalization or Prolonged Hospitalization	A0705	Battery Problem	02/12/2021	02/12/2021
									A1107	Data Problem	02/12/2021	02/12/2021
							F12	Serious Injury/ Illness/ Impairment	A0705	Battery Problem	02/12/2021	02/12/2021
									A1107	Data Problem	02/12/2021	02/12/2021
							F1905	Device Revision or Replacement	A0705	Battery Problem	02/12/2021	02/12/2021
									A1107	Data Problem	02/12/2021	02/12/2021
7242	2022	All other reportable incidents	35223	Endocardial/interventricular septal pacing lead	CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A0401	Break		
									A070909	Oversensing		
									A072202	Low Impedance		
							F1908	Prolonged Surgery	A0401	Break		
									A070909	Oversensing		

									A072202	Low Impedance		
7243	2022	All other reportable incidents	47265	Dual-chamber implantable pacemaker, rate-responsive	CT141	Electrodes/Leads and associated devices	E0601	Arrhythmia	A0712	Pacing Problem	08/07/2022	08/07/2022
									A071205	Pacemaker Found in Back-Up Mode	08/07/2022	08/07/2022
							E2403	No Clinical Signs, Symptoms or Conditions	A0712	Pacing Problem	08/07/2022	08/07/2022
									A071205	Pacemaker Found in Back-Up Mode	08/07/2022	08/07/2022
							F1908	Prolonged Surgery	A0712	Pacing Problem	08/07/2022	08/07/2022
									A071205	Pacemaker Found in Back-Up Mode	08/07/2022	08/07/2022
					CT145	Cardiovascular devices	E0601	Arrhythmia	A0712	Pacing Problem	08/07/2022	08/07/2022
									A071205	Pacemaker Found in Back-Up Mode	08/07/2022	08/07/2022
							E2403	No Clinical Signs, Symptoms or Conditions	A0712	Pacing Problem	08/07/2022	08/07/2022
									A071205	Pacemaker Found in Back-Up Mode	08/07/2022	08/07/2022
							F1908	Prolonged Surgery	A0712	Pacing Problem	08/07/2022	08/07/2022
									A071205	Pacemaker Found in Back-Up Mode	08/07/2022	08/07/2022
					CT1684	Cardiac pacemakers and associated devices	E0601	Arrhythmia	A0712	Pacing Problem	08/07/2022	08/07/2022
									A071205	Pacemaker Found in Back-Up Mode	08/07/2022	08/07/2022
							E2403	No Clinical Signs, Symptoms or Conditions	A0712	Pacing Problem	08/07/2022	08/07/2022
									A071205	Pacemaker Found in Back-Up Mode	08/07/2022	08/07/2022
							F1908	Prolonged Surgery	A0712	Pacing Problem	08/07/2022	08/07/2022
									A071205	Pacemaker Found in Back-Up Mode	08/07/2022	08/07/2022
					CT1685	Cardiac pacemakers	E0601	Arrhythmia	A0712	Pacing Problem	08/07/2022	08/07/2022
									A071205	Pacemaker Found in Back-Up Mode	08/07/2022	08/07/2022
							E2403	No Clinical Signs, Symptoms or Conditions	A0712	Pacing Problem	08/07/2022	08/07/2022
									A071205	Pacemaker Found in Back-Up Mode	08/07/2022	08/07/2022
							F1908	Prolonged Surgery	A0712	Pacing Problem	08/07/2022	08/07/2022
									A071205	Pacemaker Found in Back-Up Mode	08/07/2022	08/07/2022
					CT233	Surgical	E0601	Arrhythmia	A0712	Pacing Problem	08/07/2022	08/07/2022
									A071205	Pacemaker Found in Back-Up Mode	08/07/2022	08/07/2022
							E2403	No Clinical Signs, Symptoms or Conditions	A0712	Pacing Problem	08/07/2022	08/07/2022
									A071205	Pacemaker Found in Back-Up Mode	08/07/2022	08/07/2022
							F1908	Prolonged Surgery	A0712	Pacing Problem	08/07/2022	08/07/2022
									A071205	Pacemaker Found in Back-Up Mode	08/07/2022	08/07/2022
					CT2406	Implantable	E0601	Arrhythmia	A0712	Pacing Problem	08/07/2022	08/07/2022
									A071205	Pacemaker Found in Back-Up Mode	08/07/2022	08/07/2022
							E2403	No Clinical Signs, Symptoms or Conditions	A0712	Pacing Problem	08/07/2022	08/07/2022
									A071205	Pacemaker Found in Back-Up Mode	08/07/2022	08/07/2022
							F1908	Prolonged Surgery	A0712	Pacing Problem	08/07/2022	08/07/2022
									A071205	Pacemaker Found in Back-Up Mode	08/07/2022	08/07/2022
					CT2407	Active implantable	E0601	Arrhythmia	A0712	Pacing Problem	08/07/2022	08/07/2022
									A071205	Pacemaker Found in Back-Up Mode	08/07/2022	08/07/2022
							E2403	No Clinical Signs, Symptoms or Conditions	A0712	Pacing Problem	08/07/2022	08/07/2022
									A071205	Pacemaker Found in Back-Up Mode	08/07/2022	08/07/2022
							F1908	Prolonged Surgery	A0712	Pacing Problem	08/07/2022	08/07/2022
									A071205	Pacemaker Found in Back-Up Mode	08/07/2022	08/07/2022
					CT257	Implantable cardiac pacemakers	E0601	Arrhythmia	A0712	Pacing Problem	08/07/2022	08/07/2022
									A071205	Pacemaker Found in Back-Up Mode	08/07/2022	08/07/2022
							E2403	No Clinical Signs, Symptoms or Conditions	A0712	Pacing Problem	08/07/2022	08/07/2022
									A071205	Pacemaker Found in Back-Up Mode	08/07/2022	08/07/2022
							F1908	Prolonged Surgery	A0712	Pacing Problem	08/07/2022	08/07/2022
									A071205	Pacemaker Found in Back-Up Mode	08/07/2022	08/07/2022
					CT2727	Active implant devices/ancillaries	E0601	Arrhythmia	A0712	Pacing Problem	08/07/2022	08/07/2022
									A071205	Pacemaker Found in Back-Up Mode	08/07/2022	08/07/2022
							E2403	No Clinical Signs, Symptoms or Conditions	A0712	Pacing Problem	08/07/2022	08/07/2022
									A071205	Pacemaker Found in Back-Up Mode	08/07/2022	08/07/2022
							F1908	Prolonged Surgery	A0712	Pacing Problem	08/07/2022	08/07/2022
									A071205	Pacemaker Found in Back-Up Mode	08/07/2022	08/07/2022
					CT2750	Implantable cardiac electrical stimulators and associated devices	E0601	Arrhythmia	A0712	Pacing Problem	08/07/2022	08/07/2022
									A071205	Pacemaker Found in Back-Up Mode	08/07/2022	08/07/2022
							E2403	No Clinical Signs, Symptoms or Conditions	A0712	Pacing Problem	08/07/2022	08/07/2022
									A071205	Pacemaker Found in Back-Up Mode	08/07/2022	08/07/2022
							F1908	Prolonged Surgery	A0712	Pacing Problem	08/07/2022	08/07/2022
									A071205	Pacemaker Found in Back-Up Mode	08/07/2022	08/07/2022
					CT2866	Implantable cardiac electrical stimulators	E0601	Arrhythmia	A0712	Pacing Problem	08/07/2022	08/07/2022
									A071205	Pacemaker Found in Back-Up Mode	08/07/2022	08/07/2022
							E2403	No Clinical Signs, Symptoms or Conditions	A0712	Pacing Problem	08/07/2022	08/07/2022
									A071205	Pacemaker Found in Back-Up Mode	08/07/2022	08/07/2022
							F1908	Prolonged Surgery	A0712	Pacing Problem	08/07/2022	08/07/2022
									A071205	Pacemaker Found in Back-Up Mode	08/07/2022	08/07/2022
					CT2985	Active	E0601	Arrhythmia	A0712	Pacing Problem	08/07/2022	08/07/2022
									A071205	Pacemaker Found in Back-Up Mode	08/07/2022	08/07/2022
							E2403	No Clinical Signs, Symptoms or Conditions	A0712	Pacing Problem	08/07/2022	08/07/2022
									A071205	Pacemaker Found in Back-Up Mode	08/07/2022	08/07/2022
							F1908	Prolonged Surgery	A0712	Pacing Problem	08/07/2022	08/07/2022
									A071205	Pacemaker Found in Back-Up Mode	08/07/2022	08/07/2022
					CT3001	Implantable metal	E0601	Arrhythmia	A0712	Pacing Problem	08/07/2022	08/07/2022
									A071205	Pacemaker Found in Back-Up Mode	08/07/2022	08/07/2022
							E2403	No Clinical Signs, Symptoms or Conditions	A0712	Pacing Problem	08/07/2022	08/07/2022
									A071205	Pacemaker Found in Back-Up Mode	08/07/2022	08/07/2022
							F1908	Prolonged Surgery	A0712	Pacing Problem	08/07/2022	08/07/2022
									A071205	Pacemaker Found in Back-Up Mode	08/07/2022	08/07/2022
					CT3002	Implantable synthetic polymer	E0601	Arrhythmia	A0712	Pacing Problem	08/07/2022	08/07/2022
									A071205	Pacemaker Found in Back-Up Mode	08/07/2022	08/07/2022
							E2403	No Clinical Signs, Symptoms or Conditions	A0712	Pacing Problem	08/07/2022	08/07/2022
									A071205	Pacemaker Found in Back-Up Mode	08/07/2022	08/07/2022
							F1908	Prolonged Surgery	A0712	Pacing Problem	08/07/2022	08/07/2022
									A071205	Pacemaker Found in Back-Up Mode	08/07/2022	08/07/2022
					CT3006	Implantable inorganic material	E0601	Arrhythmia	A0712	Pacing Problem	08/07/2022	08/07/2022
									A071205	Pacemaker Found in Back-Up Mode	08/07/2022	08/07/2022
							E2403	No Clinical Signs, Symptoms or Conditions	A0712	Pacing Problem	08/07/2022	08/07/2022

							F1908	Prolonged Surgery	A071205	Pacemaker Found in Back-Up Mode	08/07/2022	08/07/2022
									A0712	Pacing Problem	08/07/2022	08/07/2022
									A071205	Pacemaker Found in Back-Up Mode	08/07/2022	08/07/2022
					CT314	Electrically powered	E0601	Arrhythmia	A0712	Pacing Problem	08/07/2022	08/07/2022
									A071205	Pacemaker Found in Back-Up Mode	08/07/2022	08/07/2022
							E2403	No Clinical Signs, Symptoms or Conditions	A0712	Pacing Problem	08/07/2022	08/07/2022
									A071205	Pacemaker Found in Back-Up Mode	08/07/2022	08/07/2022
							F1908	Prolonged Surgery	A0712	Pacing Problem	08/07/2022	08/07/2022
									A071205	Pacemaker Found in Back-Up Mode	08/07/2022	08/07/2022
					CT321	Long-term surgical invasive	E0601	Arrhythmia	A0712	Pacing Problem	08/07/2022	08/07/2022
									A071205	Pacemaker Found in Back-Up Mode	08/07/2022	08/07/2022
							E2403	No Clinical Signs, Symptoms or Conditions	A0712	Pacing Problem	08/07/2022	08/07/2022
									A071205	Pacemaker Found in Back-Up Mode	08/07/2022	08/07/2022
							F1908	Prolonged Surgery	A0712	Pacing Problem	08/07/2022	08/07/2022
									A071205	Pacemaker Found in Back-Up Mode	08/07/2022	08/07/2022
					CT449	Generators and associated devices	E0601	Arrhythmia	A0712	Pacing Problem	08/07/2022	08/07/2022
									A071205	Pacemaker Found in Back-Up Mode	08/07/2022	08/07/2022
							E2403	No Clinical Signs, Symptoms or Conditions	A0712	Pacing Problem	08/07/2022	08/07/2022
									A071205	Pacemaker Found in Back-Up Mode	08/07/2022	08/07/2022
							F1908	Prolonged Surgery	A0712	Pacing Problem	08/07/2022	08/07/2022
									A071205	Pacemaker Found in Back-Up Mode	08/07/2022	08/07/2022
					CT981	Single use	E0601	Arrhythmia	A0712	Pacing Problem	08/07/2022	08/07/2022
									A071205	Pacemaker Found in Back-Up Mode	08/07/2022	08/07/2022
							E2403	No Clinical Signs, Symptoms or Conditions	A0712	Pacing Problem	08/07/2022	08/07/2022
									A071205	Pacemaker Found in Back-Up Mode	08/07/2022	08/07/2022
							F1908	Prolonged Surgery	A0712	Pacing Problem	08/07/2022	08/07/2022
									A071205	Pacemaker Found in Back-Up Mode	08/07/2022	08/07/2022
					CT983	Surgical invasive	E0601	Arrhythmia	A0712	Pacing Problem	08/07/2022	08/07/2022
									A071205	Pacemaker Found in Back-Up Mode	08/07/2022	08/07/2022
							E2403	No Clinical Signs, Symptoms or Conditions	A0712	Pacing Problem	08/07/2022	08/07/2022
									A071205	Pacemaker Found in Back-Up Mode	08/07/2022	08/07/2022
							F1908	Prolonged Surgery	A0712	Pacing Problem	08/07/2022	08/07/2022
									A071205	Pacemaker Found in Back-Up Mode	08/07/2022	08/07/2022
7244	2021	All other reportable incidents	35223	Endocardial/interventricular septal pacing lead	CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A0401	Break		
									A0701	Capturing Problem		
									A070909	Oversensing		
							F1901	Additional Surgery	A0401	Break		
									A0701	Capturing Problem		
									A070909	Oversensing		
7245	2021	All other reportable incidents	35223	Endocardial/interventricular septal pacing lead	CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070909	Oversensing		
							F1901	Additional Surgery	A070909	Oversensing		
7246	2021	All other reportable incidents	35852	Single-chamber implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A0908	Incorrect, Inadequate or Imprecise Result or Readings		
							F26	No Health Consequences or Impact	A0908	Incorrect, Inadequate or Imprecise Result or Readings		
7247	2021	All other reportable incidents	17882	Physiologic-monitoring defibrillation system	CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070908	Failure to Sense		
							F26	No Health Consequences or Impact	A070908	Failure to Sense		
7248	2022	All other reportable incidents	47270	Cardiac resynchronization therapy implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A25	No Apparent Adverse Event	20/07/2022	20/07/2022
							F02	Death	A25	No Apparent Adverse Event	20/07/2022	20/07/2022
7249	2021	All other reportable incidents	47267	Single-chamber implantable pacemaker, rate-responsive	CT141	Electrodes/Leads and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem	10/01/2022	10/01/2022
									A070504	Premature Discharge of Battery	10/01/2022	10/01/2022
							F08	Hospitalization or Prolonged Hospitalization	A0705	Battery Problem	10/01/2022	10/01/2022
									A070504	Premature Discharge of Battery	10/01/2022	10/01/2022
							F1905	Device Revision or Replacement	A0705	Battery Problem	10/01/2022	10/01/2022
									A070504	Premature Discharge of Battery	10/01/2022	10/01/2022
					CT145	Cardiovascular devices	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem	10/01/2022	10/01/2022
									A070504	Premature Discharge of Battery	10/01/2022	10/01/2022
							F08	Hospitalization or Prolonged Hospitalization	A0705	Battery Problem	10/01/2022	10/01/2022
									A070504	Premature Discharge of Battery	10/01/2022	10/01/2022
							F1905	Device Revision or Replacement	A0705	Battery Problem	10/01/2022	10/01/2022
									A070504	Premature Discharge of Battery	10/01/2022	10/01/2022
					CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem	10/01/2022	10/01/2022
									A070504	Premature Discharge of Battery	10/01/2022	10/01/2022
							F08	Hospitalization or Prolonged Hospitalization	A0705	Battery Problem	10/01/2022	10/01/2022
									A070504	Premature Discharge of Battery	10/01/2022	10/01/2022
							F1905	Device Revision or Replacement	A0705	Battery Problem	10/01/2022	10/01/2022
									A070504	Premature Discharge of Battery	10/01/2022	10/01/2022
					CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem	10/01/2022	10/01/2022
									A070504	Premature Discharge of Battery	10/01/2022	10/01/2022
							F08	Hospitalization or Prolonged Hospitalization	A0705	Battery Problem	10/01/2022	10/01/2022
									A070504	Premature Discharge of Battery	10/01/2022	10/01/2022
							F1905	Device Revision or Replacement	A0705	Battery Problem	10/01/2022	10/01/2022
									A070504	Premature Discharge of Battery	10/01/2022	10/01/2022
					CT233	Surgical	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem	10/01/2022	10/01/2022
									A070504	Premature Discharge of Battery	10/01/2022	10/01/2022
							F08	Hospitalization or Prolonged Hospitalization	A0705	Battery Problem	10/01/2022	10/01/2022
									A070504	Premature Discharge of Battery	10/01/2022	10/01/2022
							F1905	Device Revision or Replacement	A0705	Battery Problem	10/01/2022	10/01/2022
									A070504	Premature Discharge of Battery	10/01/2022	10/01/2022
					CT2406	Implantable	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem	10/01/2022	10/01/2022
									A070504	Premature Discharge of Battery	10/01/2022	10/01/2022
							F08	Hospitalization or Prolonged Hospitalization	A0705	Battery Problem	10/01/2022	10/01/2022
									A070504	Premature Discharge of Battery	10/01/2022	10/01/2022
							F1905	Device Revision or Replacement	A0705	Battery Problem	10/01/2022	10/01/2022
									A070504	Premature Discharge of Battery	10/01/2022	10/01/2022
					CT2407	Active implantable	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem	10/01/2022	10/01/2022
									A070504	Premature Discharge of Battery	10/01/2022	10/01/2022
							F08	Hospitalization or Prolonged Hospitalization	A0705	Battery Problem	10/01/2022	10/01/2022

									A070504	Premature Discharge of Battery	10/01/2022	10/01/2022
							F1905	Device Revision or Replacement	A0705	Battery Problem	10/01/2022	10/01/2022
									A070504	Premature Discharge of Battery	10/01/2022	10/01/2022
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem	10/01/2022	10/01/2022
									A070504	Premature Discharge of Battery	10/01/2022	10/01/2022
							F08	Hospitalization or Prolonged Hospitalization	A0705	Battery Problem	10/01/2022	10/01/2022
									A070504	Premature Discharge of Battery	10/01/2022	10/01/2022
							F1905	Device Revision or Replacement	A0705	Battery Problem	10/01/2022	10/01/2022
									A070504	Premature Discharge of Battery	10/01/2022	10/01/2022
					CT2727	Active implant devices/ancillaries	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem	10/01/2022	10/01/2022
									A070504	Premature Discharge of Battery	10/01/2022	10/01/2022
							F08	Hospitalization or Prolonged Hospitalization	A0705	Battery Problem	10/01/2022	10/01/2022
									A070504	Premature Discharge of Battery	10/01/2022	10/01/2022
							F1905	Device Revision or Replacement	A0705	Battery Problem	10/01/2022	10/01/2022
									A070504	Premature Discharge of Battery	10/01/2022	10/01/2022
					CT2750	Implantable cardiac electrical stimulators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem	10/01/2022	10/01/2022
									A070504	Premature Discharge of Battery	10/01/2022	10/01/2022
							F08	Hospitalization or Prolonged Hospitalization	A0705	Battery Problem	10/01/2022	10/01/2022
									A070504	Premature Discharge of Battery	10/01/2022	10/01/2022
							F1905	Device Revision or Replacement	A0705	Battery Problem	10/01/2022	10/01/2022
									A070504	Premature Discharge of Battery	10/01/2022	10/01/2022
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem	10/01/2022	10/01/2022
									A070504	Premature Discharge of Battery	10/01/2022	10/01/2022
							F08	Hospitalization or Prolonged Hospitalization	A0705	Battery Problem	10/01/2022	10/01/2022
									A070504	Premature Discharge of Battery	10/01/2022	10/01/2022
							F1905	Device Revision or Replacement	A0705	Battery Problem	10/01/2022	10/01/2022
									A070504	Premature Discharge of Battery	10/01/2022	10/01/2022
					CT2985	Active	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem	10/01/2022	10/01/2022
									A070504	Premature Discharge of Battery	10/01/2022	10/01/2022
							F08	Hospitalization or Prolonged Hospitalization	A0705	Battery Problem	10/01/2022	10/01/2022
									A070504	Premature Discharge of Battery	10/01/2022	10/01/2022
							F1905	Device Revision or Replacement	A0705	Battery Problem	10/01/2022	10/01/2022
									A070504	Premature Discharge of Battery	10/01/2022	10/01/2022
					CT3001	Implantable metal	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem	10/01/2022	10/01/2022
									A070504	Premature Discharge of Battery	10/01/2022	10/01/2022
							F08	Hospitalization or Prolonged Hospitalization	A0705	Battery Problem	10/01/2022	10/01/2022
									A070504	Premature Discharge of Battery	10/01/2022	10/01/2022
							F1905	Device Revision or Replacement	A0705	Battery Problem	10/01/2022	10/01/2022
									A070504	Premature Discharge of Battery	10/01/2022	10/01/2022
					CT3002	Implantable synthetic polymer	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem	10/01/2022	10/01/2022
									A070504	Premature Discharge of Battery	10/01/2022	10/01/2022
							F08	Hospitalization or Prolonged Hospitalization	A0705	Battery Problem	10/01/2022	10/01/2022
									A070504	Premature Discharge of Battery	10/01/2022	10/01/2022
							F1905	Device Revision or Replacement	A0705	Battery Problem	10/01/2022	10/01/2022
									A070504	Premature Discharge of Battery	10/01/2022	10/01/2022
					CT3006	Implantable inorganic material	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem	10/01/2022	10/01/2022
									A070504	Premature Discharge of Battery	10/01/2022	10/01/2022
							F08	Hospitalization or Prolonged Hospitalization	A0705	Battery Problem	10/01/2022	10/01/2022
									A070504	Premature Discharge of Battery	10/01/2022	10/01/2022
							F1905	Device Revision or Replacement	A0705	Battery Problem	10/01/2022	10/01/2022
									A070504	Premature Discharge of Battery	10/01/2022	10/01/2022
					CT314	Electrically powered	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem	10/01/2022	10/01/2022
									A070504	Premature Discharge of Battery	10/01/2022	10/01/2022
							F08	Hospitalization or Prolonged Hospitalization	A0705	Battery Problem	10/01/2022	10/01/2022
									A070504	Premature Discharge of Battery	10/01/2022	10/01/2022
							F1905	Device Revision or Replacement	A0705	Battery Problem	10/01/2022	10/01/2022
									A070504	Premature Discharge of Battery	10/01/2022	10/01/2022
					CT321	Long-term surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem	10/01/2022	10/01/2022
									A070504	Premature Discharge of Battery	10/01/2022	10/01/2022
							F08	Hospitalization or Prolonged Hospitalization	A0705	Battery Problem	10/01/2022	10/01/2022
									A070504	Premature Discharge of Battery	10/01/2022	10/01/2022
							F1905	Device Revision or Replacement	A0705	Battery Problem	10/01/2022	10/01/2022
									A070504	Premature Discharge of Battery	10/01/2022	10/01/2022
					CT449	Generators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem	10/01/2022	10/01/2022
									A070504	Premature Discharge of Battery	10/01/2022	10/01/2022
							F08	Hospitalization or Prolonged Hospitalization	A0705	Battery Problem	10/01/2022	10/01/2022
									A070504	Premature Discharge of Battery	10/01/2022	10/01/2022
							F1905	Device Revision or Replacement	A0705	Battery Problem	10/01/2022	10/01/2022
									A070504	Premature Discharge of Battery	10/01/2022	10/01/2022
					CT981	Single use	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem	10/01/2022	10/01/2022
									A070504	Premature Discharge of Battery	10/01/2022	10/01/2022
							F08	Hospitalization or Prolonged Hospitalization	A0705	Battery Problem	10/01/2022	10/01/2022
									A070504	Premature Discharge of Battery	10/01/2022	10/01/2022
							F1905	Device Revision or Replacement	A0705	Battery Problem	10/01/2022	10/01/2022
									A070504	Premature Discharge of Battery	10/01/2022	10/01/2022
					CT983	Surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem	10/01/2022	10/01/2022
									A070504	Premature Discharge of Battery	10/01/2022	10/01/2022
							F08	Hospitalization or Prolonged Hospitalization	A0705	Battery Problem	10/01/2022	10/01/2022
									A070504	Premature Discharge of Battery	10/01/2022	10/01/2022
							F1905	Device Revision or Replacement	A0705	Battery Problem	10/01/2022	10/01/2022
									A070504	Premature Discharge of Battery	10/01/2022	10/01/2022
7250	2022	All other reportable incidents	35852	Single-chamber implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A072201	High Impedance		
							F26	No Health Consequences or Impact	A072201	High Impedance		
7251	2022	All other reportable incidents	47263	Cardiac resynchronization therapy implantable pacemaker	CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A0908	Incorrect, Inadequate or Imprecise Result or Readings	13/07/2022	13/07/2022
							F08	Hospitalization or Prolonged Hospitalization	A0908	Incorrect, Inadequate or Imprecise Result or Readings	13/07/2022	13/07/2022

							F1903	Device Explantation	A0908	Incorrect, Inadequate or Imprecise Result or Readings	13/07/2022	13/07/2022
					CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A0908	Incorrect, Inadequate or Imprecise Result or Readings	13/07/2022	13/07/2022
							F08	Hospitalization or Prolonged Hospitalization	A0908	Incorrect, Inadequate or Imprecise Result or Readings	13/07/2022	13/07/2022
							F1903	Device Explantation	A0908	Incorrect, Inadequate or Imprecise Result or Readings	13/07/2022	13/07/2022
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A0908	Incorrect, Inadequate or Imprecise Result or Readings	13/07/2022	13/07/2022
							F08	Hospitalization or Prolonged Hospitalization	A0908	Incorrect, Inadequate or Imprecise Result or Readings	13/07/2022	13/07/2022
							F1903	Device Explantation	A0908	Incorrect, Inadequate or Imprecise Result or Readings	13/07/2022	13/07/2022
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A0908	Incorrect, Inadequate or Imprecise Result or Readings	13/07/2022	13/07/2022
							F08	Hospitalization or Prolonged Hospitalization	A0908	Incorrect, Inadequate or Imprecise Result or Readings	13/07/2022	13/07/2022
							F1903	Device Explantation	A0908	Incorrect, Inadequate or Imprecise Result or Readings	13/07/2022	13/07/2022
7252	2022	All other reportable incidents	35852	Single-chamber implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem	08/09/2022	08/09/2022
									A070504	Premature Discharge of Battery	08/09/2022	08/09/2022
							F08	Hospitalization or Prolonged Hospitalization	A0705	Battery Problem	08/09/2022	08/09/2022
									A070504	Premature Discharge of Battery	08/09/2022	08/09/2022
							F1905	Device Revision or Replacement	A0705	Battery Problem	08/09/2022	08/09/2022
									A070504	Premature Discharge of Battery	08/09/2022	08/09/2022
7253	2021	All other reportable incidents	35223	Endocardial/interventricular septal pacing lead	CT1684	Cardiac pacemakers and associated devices	E011903	Syncope/Fainting	A0401	Break		
									A070101	Failure to Capture		
							E2403	No Clinical Signs, Symptoms or Conditions	A0401	Break		
									A070101	Failure to Capture		
							F12	Serious Injury/ Illness/ Impairment	A0401	Break		
									A070101	Failure to Capture		
7254	2021	All other reportable incidents	35852	Single-chamber implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E2104	Electric Shock	A070909	Oversensing		
									A071302	Inappropriate/Inadequate Shock/Stimulation		
							E2403	No Clinical Signs, Symptoms or Conditions	A070909	Oversensing		
									A071302	Inappropriate/Inadequate Shock/Stimulation		
							F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A070909	Oversensing		
									A071302	Inappropriate/Inadequate Shock/Stimulation		
7255	2021	All other reportable incidents	60789	Intracardiac pacemaker	CT130	Catheters and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070103	Intermittent Capture	29/11/2021	29/11/2021
							F26	No Health Consequences or Impact	A070103	Intermittent Capture	29/11/2021	29/11/2021
					CT145	Cardiovascular devices	E2403	No Clinical Signs, Symptoms or Conditions	A070103	Intermittent Capture	29/11/2021	29/11/2021
							F26	No Health Consequences or Impact	A070103	Intermittent Capture	29/11/2021	29/11/2021
					CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070103	Intermittent Capture	29/11/2021	29/11/2021
							F26	No Health Consequences or Impact	A070103	Intermittent Capture	29/11/2021	29/11/2021
					CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A070103	Intermittent Capture	29/11/2021	29/11/2021
							F26	No Health Consequences or Impact	A070103	Intermittent Capture	29/11/2021	29/11/2021
					CT233	Surgical	E2403	No Clinical Signs, Symptoms or Conditions	A070103	Intermittent Capture	29/11/2021	29/11/2021
							F26	No Health Consequences or Impact	A070103	Intermittent Capture	29/11/2021	29/11/2021
					CT2406	Implantable	E2403	No Clinical Signs, Symptoms or Conditions	A070103	Intermittent Capture	29/11/2021	29/11/2021
							F26	No Health Consequences or Impact	A070103	Intermittent Capture	29/11/2021	29/11/2021
					CT2407	Active implantable	E2403	No Clinical Signs, Symptoms or Conditions	A070103	Intermittent Capture	29/11/2021	29/11/2021
							F26	No Health Consequences or Impact	A070103	Intermittent Capture	29/11/2021	29/11/2021
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A070103	Intermittent Capture	29/11/2021	29/11/2021
							F26	No Health Consequences or Impact	A070103	Intermittent Capture	29/11/2021	29/11/2021
					CT2727	Active implant devices/ancillaries	E2403	No Clinical Signs, Symptoms or Conditions	A070103	Intermittent Capture	29/11/2021	29/11/2021
							F26	No Health Consequences or Impact	A070103	Intermittent Capture	29/11/2021	29/11/2021
					CT2750	Implantable cardiac electrical stimulators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070103	Intermittent Capture	29/11/2021	29/11/2021
							F26	No Health Consequences or Impact	A070103	Intermittent Capture	29/11/2021	29/11/2021
					CT2818	Catheterization/Cannulation/Fluid path devices	E2403	No Clinical Signs, Symptoms or Conditions	A070103	Intermittent Capture	29/11/2021	29/11/2021
							F26	No Health Consequences or Impact	A070103	Intermittent Capture	29/11/2021	29/11/2021
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A070103	Intermittent Capture	29/11/2021	29/11/2021
							F26	No Health Consequences or Impact	A070103	Intermittent Capture	29/11/2021	29/11/2021
					CT2985	Active	E2403	No Clinical Signs, Symptoms or Conditions	A070103	Intermittent Capture	29/11/2021	29/11/2021
							F26	No Health Consequences or Impact	A070103	Intermittent Capture	29/11/2021	29/11/2021
					CT3001	Implantable metal	E2403	No Clinical Signs, Symptoms or Conditions	A070103	Intermittent Capture	29/11/2021	29/11/2021
							F26	No Health Consequences or Impact	A070103	Intermittent Capture	29/11/2021	29/11/2021
					CT3002	Implantable synthetic polymer	E2403	No Clinical Signs, Symptoms or Conditions	A070103	Intermittent Capture	29/11/2021	29/11/2021
							F26	No Health Consequences or Impact	A070103	Intermittent Capture	29/11/2021	29/11/2021
					CT3006	Implantable inorganic material	E2403	No Clinical Signs, Symptoms or Conditions	A070103	Intermittent Capture	29/11/2021	29/11/2021
							F26	No Health Consequences or Impact	A070103	Intermittent Capture	29/11/2021	29/11/2021
					CT314	Electrically powered	E2403	No Clinical Signs, Symptoms or Conditions	A070103	Intermittent Capture	29/11/2021	29/11/2021
							F26	No Health Consequences or Impact	A070103	Intermittent Capture	29/11/2021	29/11/2021
					CT321	Long-term surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A070103	Intermittent Capture	29/11/2021	29/11/2021
							F26	No Health Consequences or Impact	A070103	Intermittent Capture	29/11/2021	29/11/2021
					CT981	Single use	E2403	No Clinical Signs, Symptoms or Conditions	A070103	Intermittent Capture	29/11/2021	29/11/2021
							F26	No Health Consequences or Impact	A070103	Intermittent Capture	29/11/2021	29/11/2021
					CT983	Surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A070103	Intermittent Capture	29/11/2021	29/11/2021
							F26	No Health Consequences or Impact	A070103	Intermittent Capture	29/11/2021	29/11/2021
7256	2022	All other reportable incidents	47263	Cardiac resynchronization therapy implantable pacemaker	CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A0908	Incorrect, Inadequate or Imprecise Result or Readings	12/07/2022	12/07/2022
							F08	Hospitalization or Prolonged Hospitalization	A0908	Incorrect, Inadequate or Imprecise Result or Readings	12/07/2022	12/07/2022
							F1905	Device Revision or Replacement	A0908	Incorrect, Inadequate or Imprecise Result or Readings	12/07/2022	12/07/2022
					CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A0908	Incorrect, Inadequate or Imprecise Result or Readings	12/07/2022	12/07/2022
							F08	Hospitalization or Prolonged Hospitalization	A0908	Incorrect, Inadequate or Imprecise Result or Readings	12/07/2022	12/07/2022
							F1905	Device Revision or Replacement	A0908	Incorrect, Inadequate or Imprecise Result or Readings	12/07/2022	12/07/2022
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A0908	Incorrect, Inadequate or Imprecise Result or Readings	12/07/2022	12/07/2022

						F08	Hospitalization or Prolonged Hospitalization	A0908	Incorrect, Inadequate or Imprecise Result or Readings	12/07/2022	12/07/2022
						F1905	Device Revision or Replacement	A0908	Incorrect, Inadequate or Imprecise Result or Readings	12/07/2022	12/07/2022
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A0908	Incorrect, Inadequate or Imprecise Result or Readings	12/07/2022	12/07/2022
							F08	Hospitalization or Prolonged Hospitalization	A0908	Incorrect, Inadequate or Imprecise Result or Readings	12/07/2022	12/07/2022
							F1905	Device Revision or Replacement	A0908	Incorrect, Inadequate or Imprecise Result or Readings	12/07/2022	12/07/2022
7257	2022	All other reportable incidents	35852	Single-chamber implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	15/08/2022	15/08/2022
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	15/08/2022	15/08/2022
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	15/08/2022	15/08/2022
7258	2021	All other reportable incidents	16995	Intracardiac electrode	CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A0501	Detachment of Device or Device Component	03/11/2021	03/11/2021
							F27	Problem Identified During Non-Clinical Procedure	A0501	Detachment of Device or Device Component	03/11/2021	03/11/2021
7259	2022	All other reportable incidents	47265	Dual-chamber implantable pacemaker, rate-responsive	CT141	Electrodes/Leads and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT145	Cardiovascular devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT233	Surgical	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT2406	Implantable	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT2407	Active implantable	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT2727	Active implant devices/ancillaries	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT2750	Implantable cardiac electrical stimulators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT2985	Active	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT3001	Implantable metal	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT3002	Implantable synthetic polymer	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT3006	Implantable inorganic material	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT314	Electrically powered	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT321	Long-term surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT449	Generators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT981	Single use	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT983	Surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
7260	2021	All other reportable incidents	47265	Dual-chamber implantable pacemaker, rate-responsive	CT141	Electrodes/Leads and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
					CT145	Cardiovascular devices	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
					CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
					CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
					CT233	Surgical	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
					CT2406	Implantable	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
					CT2407	Active implantable	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
					CT2727	Active implant devices/ancillaries	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		

							F26	No Health Consequences or Impact	A0705	Battery Problem			
									A070504	Premature Discharge of Battery			
					CT2750	Implantable cardiac electrical stimulators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem			
									A070504	Premature Discharge of Battery			
							F26	No Health Consequences or Impact	A0705	Battery Problem			
									A070504	Premature Discharge of Battery			
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem			
									A070504	Premature Discharge of Battery			
							F26	No Health Consequences or Impact	A0705	Battery Problem			
									A070504	Premature Discharge of Battery			
					CT2985	Active	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem			
									A070504	Premature Discharge of Battery			
							F26	No Health Consequences or Impact	A0705	Battery Problem			
									A070504	Premature Discharge of Battery			
					CT3001	Implantable metal	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem			
									A070504	Premature Discharge of Battery			
							F26	No Health Consequences or Impact	A0705	Battery Problem			
									A070504	Premature Discharge of Battery			
					CT3002	Implantable synthetic polymer	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem			
									A070504	Premature Discharge of Battery			
							F26	No Health Consequences or Impact	A0705	Battery Problem			
									A070504	Premature Discharge of Battery			
					CT3006	Implantable inorganic material	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem			
									A070504	Premature Discharge of Battery			
							F26	No Health Consequences or Impact	A0705	Battery Problem			
									A070504	Premature Discharge of Battery			
					CT314	Electrically powered	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem			
									A070504	Premature Discharge of Battery			
							F26	No Health Consequences or Impact	A0705	Battery Problem			
									A070504	Premature Discharge of Battery			
					CT321	Long-term surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem			
									A070504	Premature Discharge of Battery			
							F26	No Health Consequences or Impact	A0705	Battery Problem			
									A070504	Premature Discharge of Battery			
					CT449	Generators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem			
									A070504	Premature Discharge of Battery			
							F26	No Health Consequences or Impact	A0705	Battery Problem			
									A070504	Premature Discharge of Battery			
					CT981	Single use	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem			
									A070504	Premature Discharge of Battery			
							F26	No Health Consequences or Impact	A0705	Battery Problem			
									A070504	Premature Discharge of Battery			
					CT983	Surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem			
									A070504	Premature Discharge of Battery			
							F26	No Health Consequences or Impact	A0705	Battery Problem			
									A070504	Premature Discharge of Battery			
7261	2022	All other reportable incidents	35223	Endocardial/interventricular septal pacing lead	CT1684	Cardiac pacemakers and associated devices	E0506	Hemorrhage/Blood Loss/Bleeding	A070101	Failure to Capture			
									A070910	Undersensing			
							E0604	Cardiac Perforation	A070101	Failure to Capture			
									A070910	Undersensing			
							E0619	Pericardial Effusion	A070101	Failure to Capture			
									A070910	Undersensing			
							E0717	Dyspnea	A070101	Failure to Capture			
									A070910	Undersensing			
							E0722	Hemothorax	A070101	Failure to Capture			
									A070910	Undersensing			
							E233001	Chest Pain	A070101	Failure to Capture			
									A070910	Undersensing			
							E2403	No Clinical Signs, Symptoms or Conditions	A070101	Failure to Capture			
									A070910	Undersensing			
							F1203	Life Threatening Illness or Injury	A070101	Failure to Capture			
									A070910	Undersensing			
							F1905	Device Revision or Replacement	A070101	Failure to Capture			
									A070910	Undersensing			
							F2203	Imaging Required	A070101	Failure to Capture			
									A070910	Undersensing			
							F23	Unexpected Medical Intervention	A070101	Failure to Capture			
									A070910	Undersensing			
7262	2022	All other reportable incidents	35852	Single-chamber implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	15/07/2022	15/07/2022	
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	15/07/2022	15/07/2022	
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	15/07/2022	15/07/2022	
7263	2022	All other reportable incidents	35852	Single-chamber implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem	08/09/2022	08/09/2022	
									A070504	Premature Discharge of Battery	08/09/2022	08/09/2022	
							F08	Hospitalization or Prolonged Hospitalization	A0705	Battery Problem	08/09/2022	08/09/2022	
									A070504	Premature Discharge of Battery	08/09/2022	08/09/2022	
							F1903	Device Explantation	A0705	Battery Problem	08/09/2022	08/09/2022	
									A070504	Premature Discharge of Battery	08/09/2022	08/09/2022	
7264	2022	All other reportable incidents	47263	Cardiac resynchronization therapy implantable pacemaker	CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	20/07/2022	20/07/2022	
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	20/07/2022	20/07/2022	
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	20/07/2022	20/07/2022	
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery	20/07/2022	20/07/2022	
					CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	20/07/2022	20/07/2022	
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	20/07/2022	20/07/2022	
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	20/07/2022	20/07/2022	
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery	20/07/2022	20/07/2022	
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	20/07/2022	20/07/2022	
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	20/07/2022	20/07/2022	
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	20/07/2022	20/07/2022	

						F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery	20/07/2022	20/07/2022
				CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	20/07/2022	20/07/2022
						F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	20/07/2022	20/07/2022
						F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	20/07/2022	20/07/2022
						F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery	20/07/2022	20/07/2022
7265	2021	All other reportable incidents	35223	Endocardial/interventricular septal pacing lead	CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A0401	Break		
									A070101	Failure to Capture		
									A070909	Oversensing		
									A070910	Undersensing		
									A072202	Low Impedance		
							F1901	Additional Surgery	A0401	Break		
									A070101	Failure to Capture		
									A070909	Oversensing		
									A070910	Undersensing		
									A072202	Low Impedance		
7266	2022	All other reportable incidents	47265	Dual-chamber implantable pacemaker, rate-responsive	CT141	Electrodes/Leads and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A0701	Capturing Problem		
							F26	No Health Consequences or Impact	A0701	Capturing Problem		
					CT145	Cardiovascular devices	E2403	No Clinical Signs, Symptoms or Conditions	A0701	Capturing Problem		
							F26	No Health Consequences or Impact	A0701	Capturing Problem		
					CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A0701	Capturing Problem		
							F26	No Health Consequences or Impact	A0701	Capturing Problem		
					CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A0701	Capturing Problem		
							F26	No Health Consequences or Impact	A0701	Capturing Problem		
					CT233	Surgical	E2403	No Clinical Signs, Symptoms or Conditions	A0701	Capturing Problem		
							F26	No Health Consequences or Impact	A0701	Capturing Problem		
					CT2406	Implantable	E2403	No Clinical Signs, Symptoms or Conditions	A0701	Capturing Problem		
							F26	No Health Consequences or Impact	A0701	Capturing Problem		
					CT2407	Active implantable	E2403	No Clinical Signs, Symptoms or Conditions	A0701	Capturing Problem		
							F26	No Health Consequences or Impact	A0701	Capturing Problem		
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A0701	Capturing Problem		
							F26	No Health Consequences or Impact	A0701	Capturing Problem		
					CT2727	Active implant devices/ancillaries	E2403	No Clinical Signs, Symptoms or Conditions	A0701	Capturing Problem		
							F26	No Health Consequences or Impact	A0701	Capturing Problem		
					CT2750	Implantable cardiac electrical stimulators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A0701	Capturing Problem		
							F26	No Health Consequences or Impact	A0701	Capturing Problem		
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A0701	Capturing Problem		
							F26	No Health Consequences or Impact	A0701	Capturing Problem		
					CT2985	Active	E2403	No Clinical Signs, Symptoms or Conditions	A0701	Capturing Problem		
							F26	No Health Consequences or Impact	A0701	Capturing Problem		
					CT3001	Implantable metal	E2403	No Clinical Signs, Symptoms or Conditions	A0701	Capturing Problem		
							F26	No Health Consequences or Impact	A0701	Capturing Problem		
					CT3002	Implantable synthetic polymer	E2403	No Clinical Signs, Symptoms or Conditions	A0701	Capturing Problem		
							F26	No Health Consequences or Impact	A0701	Capturing Problem		
					CT3006	Implantable inorganic material	E2403	No Clinical Signs, Symptoms or Conditions	A0701	Capturing Problem		
							F26	No Health Consequences or Impact	A0701	Capturing Problem		
					CT314	Electrically powered	E2403	No Clinical Signs, Symptoms or Conditions	A0701	Capturing Problem		
							F26	No Health Consequences or Impact	A0701	Capturing Problem		
					CT321	Long-term surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A0701	Capturing Problem		
							F26	No Health Consequences or Impact	A0701	Capturing Problem		
					CT449	Generators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A0701	Capturing Problem		
							F26	No Health Consequences or Impact	A0701	Capturing Problem		
					CT981	Single use	E2403	No Clinical Signs, Symptoms or Conditions	A0701	Capturing Problem		
							F26	No Health Consequences or Impact	A0701	Capturing Problem		
					CT983	Surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A0701	Capturing Problem		
							F26	No Health Consequences or Impact	A0701	Capturing Problem		
7267	2021	All other reportable incidents	47265	Dual-chamber implantable pacemaker, rate-responsive	CT141	Electrodes/Leads and associated devices	E0112	Dizziness	A0712	Pacing Problem		
							E060109	Tachycardia	A0712	Pacing Problem		
							E0618	Non Specific EKG/ECG Changes	A0712	Pacing Problem		
							E2403	No Clinical Signs, Symptoms or Conditions	A0712	Pacing Problem		
							F12	Serious Injury/ Illness/ Impairment	A0712	Pacing Problem		
					CT145	Cardiovascular devices	E0112	Dizziness	A0712	Pacing Problem		
							E060109	Tachycardia	A0712	Pacing Problem		
							E0618	Non Specific EKG/ECG Changes	A0712	Pacing Problem		
							E2403	No Clinical Signs, Symptoms or Conditions	A0712	Pacing Problem		
							F12	Serious Injury/ Illness/ Impairment	A0712	Pacing Problem		
					CT1684	Cardiac pacemakers and associated devices	E0112	Dizziness	A0712	Pacing Problem		
							E060109	Tachycardia	A0712	Pacing Problem		
							E0618	Non Specific EKG/ECG Changes	A0712	Pacing Problem		
							E2403	No Clinical Signs, Symptoms or Conditions	A0712	Pacing Problem		
							F12	Serious Injury/ Illness/ Impairment	A0712	Pacing Problem		
					CT1685	Cardiac pacemakers	E0112	Dizziness	A0712	Pacing Problem		
							E060109	Tachycardia	A0712	Pacing Problem		
							E0618	Non Specific EKG/ECG Changes	A0712	Pacing Problem		
							E2403	No Clinical Signs, Symptoms or Conditions	A0712	Pacing Problem		
							F12	Serious Injury/ Illness/ Impairment	A0712	Pacing Problem		
					CT233	Surgical	E0112	Dizziness	A0712	Pacing Problem		
							E060109	Tachycardia	A0712	Pacing Problem		
							E0618	Non Specific EKG/ECG Changes	A0712	Pacing Problem		
							E2403	No Clinical Signs, Symptoms or Conditions	A0712	Pacing Problem		
							F12	Serious Injury/ Illness/ Impairment	A0712	Pacing Problem		
					CT2406	Implantable	E0112	Dizziness	A0712	Pacing Problem		
							E060109	Tachycardia	A0712	Pacing Problem		
							E0618	Non Specific EKG/ECG Changes	A0712	Pacing Problem		
							E2403	No Clinical Signs, Symptoms or Conditions	A0712	Pacing Problem		
							F12	Serious Injury/ Illness/ Impairment	A0712	Pacing Problem		
					CT2407	Active implantable	E0112	Dizziness	A0712	Pacing Problem		
							E060109	Tachycardia	A0712	Pacing Problem		
							E0618	Non Specific EKG/ECG Changes	A0712	Pacing Problem		

							E2403	No Clinical Signs, Symptoms or Conditions	A0712	Pacing Problem		
							F12	Serious Injury/ Illness/ Impairment	A0712	Pacing Problem		
					CT257	Implantable cardiac pacemakers	E0112	Dizziness	A0712	Pacing Problem		
							E060109	Tachycardia	A0712	Pacing Problem		
							E0618	Non Specific EKG/ECG Changes	A0712	Pacing Problem		
							E2403	No Clinical Signs, Symptoms or Conditions	A0712	Pacing Problem		
							F12	Serious Injury/ Illness/ Impairment	A0712	Pacing Problem		
					CT2727	Active implant devices/ancillaries	E0112	Dizziness	A0712	Pacing Problem		
							E060109	Tachycardia	A0712	Pacing Problem		
							E0618	Non Specific EKG/ECG Changes	A0712	Pacing Problem		
							E2403	No Clinical Signs, Symptoms or Conditions	A0712	Pacing Problem		
							F12	Serious Injury/ Illness/ Impairment	A0712	Pacing Problem		
					CT2750	Implantable cardiac electrical stimulators and associated devices	E0112	Dizziness	A0712	Pacing Problem		
							E060109	Tachycardia	A0712	Pacing Problem		
							E0618	Non Specific EKG/ECG Changes	A0712	Pacing Problem		
							E2403	No Clinical Signs, Symptoms or Conditions	A0712	Pacing Problem		
							F12	Serious Injury/ Illness/ Impairment	A0712	Pacing Problem		
					CT2866	Implantable cardiac electrical stimulators	E0112	Dizziness	A0712	Pacing Problem		
							E060109	Tachycardia	A0712	Pacing Problem		
							E0618	Non Specific EKG/ECG Changes	A0712	Pacing Problem		
							E2403	No Clinical Signs, Symptoms or Conditions	A0712	Pacing Problem		
							F12	Serious Injury/ Illness/ Impairment	A0712	Pacing Problem		
					CT2985	Active	E0112	Dizziness	A0712	Pacing Problem		
							E060109	Tachycardia	A0712	Pacing Problem		
							E0618	Non Specific EKG/ECG Changes	A0712	Pacing Problem		
							E2403	No Clinical Signs, Symptoms or Conditions	A0712	Pacing Problem		
							F12	Serious Injury/ Illness/ Impairment	A0712	Pacing Problem		
					CT3001	Implantable metal	E0112	Dizziness	A0712	Pacing Problem		
							E060109	Tachycardia	A0712	Pacing Problem		
							E0618	Non Specific EKG/ECG Changes	A0712	Pacing Problem		
							E2403	No Clinical Signs, Symptoms or Conditions	A0712	Pacing Problem		
							F12	Serious Injury/ Illness/ Impairment	A0712	Pacing Problem		
					CT3002	Implantable synthetic polymer	E0112	Dizziness	A0712	Pacing Problem		
							E060109	Tachycardia	A0712	Pacing Problem		
							E0618	Non Specific EKG/ECG Changes	A0712	Pacing Problem		
							E2403	No Clinical Signs, Symptoms or Conditions	A0712	Pacing Problem		
							F12	Serious Injury/ Illness/ Impairment	A0712	Pacing Problem		
					CT3006	Implantable inorganic material	E0112	Dizziness	A0712	Pacing Problem		
							E060109	Tachycardia	A0712	Pacing Problem		
							E0618	Non Specific EKG/ECG Changes	A0712	Pacing Problem		
							E2403	No Clinical Signs, Symptoms or Conditions	A0712	Pacing Problem		
							F12	Serious Injury/ Illness/ Impairment	A0712	Pacing Problem		
					CT314	Electrically powered	E0112	Dizziness	A0712	Pacing Problem		
							E060109	Tachycardia	A0712	Pacing Problem		
							E0618	Non Specific EKG/ECG Changes	A0712	Pacing Problem		
							E2403	No Clinical Signs, Symptoms or Conditions	A0712	Pacing Problem		
							F12	Serious Injury/ Illness/ Impairment	A0712	Pacing Problem		
					CT321	Long-term surgical invasive	E0112	Dizziness	A0712	Pacing Problem		
							E060109	Tachycardia	A0712	Pacing Problem		
							E0618	Non Specific EKG/ECG Changes	A0712	Pacing Problem		
							E2403	No Clinical Signs, Symptoms or Conditions	A0712	Pacing Problem		
							F12	Serious Injury/ Illness/ Impairment	A0712	Pacing Problem		
					CT449	Generators and associated devices	E0112	Dizziness	A0712	Pacing Problem		
							E060109	Tachycardia	A0712	Pacing Problem		
							E0618	Non Specific EKG/ECG Changes	A0712	Pacing Problem		
							E2403	No Clinical Signs, Symptoms or Conditions	A0712	Pacing Problem		
							F12	Serious Injury/ Illness/ Impairment	A0712	Pacing Problem		
					CT981	Single use	E0112	Dizziness	A0712	Pacing Problem		
							E060109	Tachycardia	A0712	Pacing Problem		
							E0618	Non Specific EKG/ECG Changes	A0712	Pacing Problem		
							E2403	No Clinical Signs, Symptoms or Conditions	A0712	Pacing Problem		
							F12	Serious Injury/ Illness/ Impairment	A0712	Pacing Problem		
					CT983	Surgical invasive	E0112	Dizziness	A0712	Pacing Problem		
							E060109	Tachycardia	A0712	Pacing Problem		
							E0618	Non Specific EKG/ECG Changes	A0712	Pacing Problem		
							E2403	No Clinical Signs, Symptoms or Conditions	A0712	Pacing Problem		
							F12	Serious Injury/ Illness/ Impairment	A0712	Pacing Problem		
7268	2021	All other reportable incidents	47265	Dual-chamber implantable pacemaker, rate-responsive	CT141	Electrodes/Leads and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
					CT145	Cardiovascular devices	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
					CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
					CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
					CT233	Surgical	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
					CT2406	Implantable	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem		

									A070504	Premature Discharge of Battery		
						F26	No Health Consequences or Impact	A0705	Battery Problem			
								A070504	Premature Discharge of Battery			
					CT2407	Active implantable	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
					CT2727	Active implant devices/ancillaries	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
					CT2750	Implantable cardiac electrical stimulators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
					CT2985	Active	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
					CT3001	Implantable metal	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
					CT3002	Implantable synthetic polymer	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
					CT3006	Implantable inorganic material	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
					CT314	Electrically powered	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
					CT321	Long-term surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
					CT449	Generators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
					CT981	Single use	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
					CT983	Surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
7269	2021	All other reportable incidents	47265	Dual-chamber implantable pacemaker, rate-responsive	CT141	Electrodes/Leads and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	04/11/2021	04/11/2021
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	04/11/2021	04/11/2021
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	04/11/2021	04/11/2021
					CT145	Cardiovascular devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	04/11/2021	04/11/2021
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	04/11/2021	04/11/2021
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	04/11/2021	04/11/2021
					CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	04/11/2021	04/11/2021
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	04/11/2021	04/11/2021
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	04/11/2021	04/11/2021
					CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	04/11/2021	04/11/2021
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	04/11/2021	04/11/2021
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	04/11/2021	04/11/2021
					CT233	Surgical	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	04/11/2021	04/11/2021
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	04/11/2021	04/11/2021
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	04/11/2021	04/11/2021
					CT2406	Implantable	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	04/11/2021	04/11/2021
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	04/11/2021	04/11/2021
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	04/11/2021	04/11/2021
					CT2407	Active implantable	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	04/11/2021	04/11/2021
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	04/11/2021	04/11/2021
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	04/11/2021	04/11/2021
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	04/11/2021	04/11/2021
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	04/11/2021	04/11/2021
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	04/11/2021	04/11/2021
					CT2727	Active implant devices/ancillaries	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	04/11/2021	04/11/2021
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	04/11/2021	04/11/2021
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	04/11/2021	04/11/2021
					CT2750	Implantable cardiac electrical stimulators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	04/11/2021	04/11/2021
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	04/11/2021	04/11/2021

							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	04/11/2021	04/11/2021
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	04/11/2021	04/11/2021
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	04/11/2021	04/11/2021
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	04/11/2021	04/11/2021
					CT2985	Active	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	04/11/2021	04/11/2021
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	04/11/2021	04/11/2021
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	04/11/2021	04/11/2021
					CT3001	Implantable metal	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	04/11/2021	04/11/2021
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	04/11/2021	04/11/2021
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	04/11/2021	04/11/2021
					CT3002	Implantable synthetic polymer	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	04/11/2021	04/11/2021
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	04/11/2021	04/11/2021
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	04/11/2021	04/11/2021
					CT3006	Implantable inorganic material	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	04/11/2021	04/11/2021
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	04/11/2021	04/11/2021
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	04/11/2021	04/11/2021
					CT314	Electrically powered	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	04/11/2021	04/11/2021
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	04/11/2021	04/11/2021
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	04/11/2021	04/11/2021
					CT321	Long-term surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	04/11/2021	04/11/2021
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	04/11/2021	04/11/2021
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	04/11/2021	04/11/2021
					CT449	Generators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	04/11/2021	04/11/2021
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	04/11/2021	04/11/2021
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	04/11/2021	04/11/2021
					CT981	Single use	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	04/11/2021	04/11/2021
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	04/11/2021	04/11/2021
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	04/11/2021	04/11/2021
					CT983	Surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	04/11/2021	04/11/2021
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	04/11/2021	04/11/2021
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	04/11/2021	04/11/2021
7270	2022	All other reportable incidents	63188	Left-ventricular wireless pacing system sensor/transmitter	CT141	Electrodes/Leads and associated devices	E0731	Pleural Effusion	A070801	Complete Loss of Power		
							E2403	No Clinical Signs, Symptoms or Conditions	A070801	Complete Loss of Power		
							F1903	Device Explantation	A070801	Complete Loss of Power		
					CT145	Cardiovascular devices	E0731	Pleural Effusion	A070801	Complete Loss of Power		
							E2403	No Clinical Signs, Symptoms or Conditions	A070801	Complete Loss of Power		
							F1903	Device Explantation	A070801	Complete Loss of Power		
					CT1577	Wireless monitoring system transmitters	E0731	Pleural Effusion	A070801	Complete Loss of Power		
							E2403	No Clinical Signs, Symptoms or Conditions	A070801	Complete Loss of Power		
							F1903	Device Explantation	A070801	Complete Loss of Power		
					CT1684	Cardiac pacemakers and associated devices	E0731	Pleural Effusion	A070801	Complete Loss of Power		
							E2403	No Clinical Signs, Symptoms or Conditions	A070801	Complete Loss of Power		
							F1903	Device Explantation	A070801	Complete Loss of Power		
					CT2317	Transmitters/Receivers/Transceivers	E0731	Pleural Effusion	A070801	Complete Loss of Power		
							E2403	No Clinical Signs, Symptoms or Conditions	A070801	Complete Loss of Power		
							F1903	Device Explantation	A070801	Complete Loss of Power		
					CT233	Surgical	E0731	Pleural Effusion	A070801	Complete Loss of Power		
							E2403	No Clinical Signs, Symptoms or Conditions	A070801	Complete Loss of Power		
							F1903	Device Explantation	A070801	Complete Loss of Power		
					CT2406	Implantable	E0731	Pleural Effusion	A070801	Complete Loss of Power		
							E2403	No Clinical Signs, Symptoms or Conditions	A070801	Complete Loss of Power		
							F1903	Device Explantation	A070801	Complete Loss of Power		
					CT2407	Active implantable	E0731	Pleural Effusion	A070801	Complete Loss of Power		
							E2403	No Clinical Signs, Symptoms or Conditions	A070801	Complete Loss of Power		
							F1903	Device Explantation	A070801	Complete Loss of Power		
					CT255	Transducers and associated devices	E0731	Pleural Effusion	A070801	Complete Loss of Power		
							E2403	No Clinical Signs, Symptoms or Conditions	A070801	Complete Loss of Power		
							F1903	Device Explantation	A070801	Complete Loss of Power		
					CT2726	Physiological monitoring/measuring devices	E0731	Pleural Effusion	A070801	Complete Loss of Power		
							E2403	No Clinical Signs, Symptoms or Conditions	A070801	Complete Loss of Power		
							F1903	Device Explantation	A070801	Complete Loss of Power		
					CT2727	Active implant devices/ancillaries	E0731	Pleural Effusion	A070801	Complete Loss of Power		
							E2403	No Clinical Signs, Symptoms or Conditions	A070801	Complete Loss of Power		
							F1903	Device Explantation	A070801	Complete Loss of Power		
					CT2750	Implantable cardiac electrical stimulators and associated devices	E0731	Pleural Effusion	A070801	Complete Loss of Power		
							E2403	No Clinical Signs, Symptoms or Conditions	A070801	Complete Loss of Power		
							F1903	Device Explantation	A070801	Complete Loss of Power		
					CT2885	Physiological monitoring/measuring system transmitters/receivers/transceivers	E0731	Pleural Effusion	A070801	Complete Loss of Power		
							E2403	No Clinical Signs, Symptoms or Conditions	A070801	Complete Loss of Power		
							F1903	Device Explantation	A070801	Complete Loss of Power		
					CT2985	Active	E0731	Pleural Effusion	A070801	Complete Loss of Power		
							E2403	No Clinical Signs, Symptoms or Conditions	A070801	Complete Loss of Power		
							F1903	Device Explantation	A070801	Complete Loss of Power		
					CT3001	Implantable metal	E0731	Pleural Effusion	A070801	Complete Loss of Power		
							E2403	No Clinical Signs, Symptoms or Conditions	A070801	Complete Loss of Power		
							F1903	Device Explantation	A070801	Complete Loss of Power		
					CT3002	Implantable synthetic polymer	E0731	Pleural Effusion	A070801	Complete Loss of Power		
							E2403	No Clinical Signs, Symptoms or Conditions	A070801	Complete Loss of Power		
							F1903	Device Explantation	A070801	Complete Loss of Power		
					CT3006	Implantable inorganic material	E0731	Pleural Effusion	A070801	Complete Loss of Power		
							E2403	No Clinical Signs, Symptoms or Conditions	A070801	Complete Loss of Power		
							F1903	Device Explantation	A070801	Complete Loss of Power		
					CT314	Electrically powered	E0731	Pleural Effusion	A070801	Complete Loss of Power		
							E2403	No Clinical Signs, Symptoms or Conditions	A070801	Complete Loss of Power		
							F1903	Device Explantation	A070801	Complete Loss of Power		
					CT321	Long-term surgical invasive	E0731	Pleural Effusion	A070801	Complete Loss of Power		
							E2403	No Clinical Signs, Symptoms or Conditions	A070801	Complete Loss of Power		
							F1903	Device Explantation	A070801	Complete Loss of Power		

					CT384	Sensors and associated devices	E0731	Pleural Effusion	A070801	Complete Loss of Power			
							E2403	No Clinical Signs, Symptoms or Conditions	A070801	Complete Loss of Power			
							F1903	Device Explantation	A070801	Complete Loss of Power			
					CT449	Generators and associated devices	E0731	Pleural Effusion	A070801	Complete Loss of Power			
							E2403	No Clinical Signs, Symptoms or Conditions	A070801	Complete Loss of Power			
							F1903	Device Explantation	A070801	Complete Loss of Power			
					CT981	Single use	E0731	Pleural Effusion	A070801	Complete Loss of Power			
							E2403	No Clinical Signs, Symptoms or Conditions	A070801	Complete Loss of Power			
							F1903	Device Explantation	A070801	Complete Loss of Power			
					CT983	Surgical invasive	E0731	Pleural Effusion	A070801	Complete Loss of Power			
							E2403	No Clinical Signs, Symptoms or Conditions	A070801	Complete Loss of Power			
							F1903	Device Explantation	A070801	Complete Loss of Power			
					CT994	Sensors	E0731	Pleural Effusion	A070801	Complete Loss of Power			
							E2403	No Clinical Signs, Symptoms or Conditions	A070801	Complete Loss of Power			
							F1903	Device Explantation	A070801	Complete Loss of Power			
7271	2022	All other reportable incidents	17882	Physiologic-monitoring defibrillation system	CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070803	Failure to Power Up			
							F27	Problem Identified During Non-Clinical Procedure	A070803	Failure to Power Up			
7272	2022	All other reportable incidents	35852	Single-chamber implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery			
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery			
7273	2021	All other reportable incidents	36241	Pacemaker/defibrillator lead	CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070909	Oversensing	19/11/2021	19/11/2021	
							F08	Hospitalization or Prolonged Hospitalization	A070909	Oversensing	19/11/2021	19/11/2021	
							F1905	Device Revision or Replacement	A070909	Oversensing	19/11/2021	19/11/2021	
7274	2021	All other reportable incidents	47263	Cardiac resynchronization therapy implantable pacemaker	CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A0908	Incorrect, Inadequate or Imprecise Result or Readings	30/11/2021	30/11/2021	
							F08	Hospitalization or Prolonged Hospitalization	A0908	Incorrect, Inadequate or Imprecise Result or Readings	30/11/2021	30/11/2021	
							F1905	Device Revision or Replacement	A0908	Incorrect, Inadequate or Imprecise Result or Readings	30/11/2021	30/11/2021	
					CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A0908	Incorrect, Inadequate or Imprecise Result or Readings	30/11/2021	30/11/2021	
							F08	Hospitalization or Prolonged Hospitalization	A0908	Incorrect, Inadequate or Imprecise Result or Readings	30/11/2021	30/11/2021	
							F1905	Device Revision or Replacement	A0908	Incorrect, Inadequate or Imprecise Result or Readings	30/11/2021	30/11/2021	
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A0908	Incorrect, Inadequate or Imprecise Result or Readings	30/11/2021	30/11/2021	
							F08	Hospitalization or Prolonged Hospitalization	A0908	Incorrect, Inadequate or Imprecise Result or Readings	30/11/2021	30/11/2021	
							F1905	Device Revision or Replacement	A0908	Incorrect, Inadequate or Imprecise Result or Readings	30/11/2021	30/11/2021	
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A0908	Incorrect, Inadequate or Imprecise Result or Readings	30/11/2021	30/11/2021	
							F08	Hospitalization or Prolonged Hospitalization	A0908	Incorrect, Inadequate or Imprecise Result or Readings	30/11/2021	30/11/2021	
							F1905	Device Revision or Replacement	A0908	Incorrect, Inadequate or Imprecise Result or Readings	30/11/2021	30/11/2021	
7275	2021	All other reportable incidents	47267	Single-chamber implantable pacemaker, rate-responsive	CT141	Electrodes/Leads and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	06/12/2021	06/12/2021	
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	06/12/2021	06/12/2021	
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	06/12/2021	06/12/2021	
					CT145	Cardiovascular devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	06/12/2021	06/12/2021	
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	06/12/2021	06/12/2021	
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	06/12/2021	06/12/2021	
					CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	06/12/2021	06/12/2021	
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	06/12/2021	06/12/2021	
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	06/12/2021	06/12/2021	
					CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	06/12/2021	06/12/2021	
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	06/12/2021	06/12/2021	
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	06/12/2021	06/12/2021	
					CT233	Surgical	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	06/12/2021	06/12/2021	
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	06/12/2021	06/12/2021	
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	06/12/2021	06/12/2021	
					CT2406	Implantable	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	06/12/2021	06/12/2021	
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	06/12/2021	06/12/2021	
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	06/12/2021	06/12/2021	
					CT2407	Active implantable	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	06/12/2021	06/12/2021	
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	06/12/2021	06/12/2021	
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	06/12/2021	06/12/2021	
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	06/12/2021	06/12/2021	
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	06/12/2021	06/12/2021	
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	06/12/2021	06/12/2021	
					CT2727	Active implant devices/ancillaries	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	06/12/2021	06/12/2021	
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	06/12/2021	06/12/2021	
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	06/12/2021	06/12/2021	
					CT2750	Implantable cardiac electrical stimulators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	06/12/2021	06/12/2021	
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	06/12/2021	06/12/2021	
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	06/12/2021	06/12/2021	
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	06/12/2021	06/12/2021	
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	06/12/2021	06/12/2021	
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	06/12/2021	06/12/2021	
					CT2985	Active	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	06/12/2021	06/12/2021	
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	06/12/2021	06/12/2021	
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	06/12/2021	06/12/2021	
					CT3001	Implantable metal	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	06/12/2021	06/12/2021	
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	06/12/2021	06/12/2021	
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	06/12/2021	06/12/2021	
					CT3002	Implantable synthetic polymer	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	06/12/2021	06/12/2021	
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	06/12/2021	06/12/2021	
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	06/12/2021	06/12/2021	
					CT3006	Implantable inorganic material	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	06/12/2021	06/12/2021	
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	06/12/2021	06/12/2021	
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	06/12/2021	06/12/2021	

					CT314	Electrically powered	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	06/12/2021	06/12/2021
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	06/12/2021	06/12/2021
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	06/12/2021	06/12/2021
					CT321	Long-term surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	06/12/2021	06/12/2021
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	06/12/2021	06/12/2021
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	06/12/2021	06/12/2021
					CT449	Generators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	06/12/2021	06/12/2021
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	06/12/2021	06/12/2021
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	06/12/2021	06/12/2021
					CT981	Single use	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	06/12/2021	06/12/2021
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	06/12/2021	06/12/2021
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	06/12/2021	06/12/2021
					CT983	Surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	06/12/2021	06/12/2021
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	06/12/2021	06/12/2021
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	06/12/2021	06/12/2021
7276	2022	All other reportable incidents	36241	Pacemaker/defibrillator lead	CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070909	Oversensing		
									A0712	Pacing Problem		
							F26	No Health Consequences or Impact	A070909	Oversensing		
									A0712	Pacing Problem		
7277	2021	All other reportable incidents	35852	Single-chamber implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E2104	Electric Shock	A070909	Oversensing		
									A071302	Inappropriate/Inadequate Shock/Stimulation		
									A1301	Failure to Read Input Signal		
							E2403	No Clinical Signs, Symptoms or Conditions	A070909	Oversensing		
									A071302	Inappropriate/Inadequate Shock/Stimulation		
									A1301	Failure to Read Input Signal		
							F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A070909	Oversensing		
									A071302	Inappropriate/Inadequate Shock/Stimulation		
									A1301	Failure to Read Input Signal		
7278	2021	All other reportable incidents	35852	Single-chamber implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	08/12/2021	08/12/2021
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	08/12/2021	08/12/2021
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	08/12/2021	08/12/2021
7279	2022	All other reportable incidents	35223	Endocardial/interventricular septal pacing lead	CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A090801	Signal Artifact/Noise	12/07/2022	12/07/2022
									A1702	Device-Device Incompatibility	12/07/2022	12/07/2022
							F08	Hospitalization or Prolonged Hospitalization	A090801	Signal Artifact/Noise	12/07/2022	12/07/2022
									A1702	Device-Device Incompatibility	12/07/2022	12/07/2022
							F12	Serious Injury/ Illness/ Impairment	A090801	Signal Artifact/Noise	12/07/2022	12/07/2022
									A1702	Device-Device Incompatibility	12/07/2022	12/07/2022
							F1905	Device Revision or Replacement	A090801	Signal Artifact/Noise	12/07/2022	12/07/2022
									A1702	Device-Device Incompatibility	12/07/2022	12/07/2022
							F23	Unexpected Medical Intervention	A090801	Signal Artifact/Noise	12/07/2022	12/07/2022
									A1702	Device-Device Incompatibility	12/07/2022	12/07/2022
7280	2021	All other reportable incidents	35223	Endocardial/interventricular septal pacing lead	CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A072201	High Impedance		
							F26	No Health Consequences or Impact	A072201	High Impedance		
7281	2022	All other reportable incidents	35223	Endocardial/interventricular septal pacing lead	CT1684	Cardiac pacemakers and associated devices	E0611	Heart Failure/Congestive Heart Failure	A051201	Device Dislodged or Dislocated	08/07/2022	08/07/2022
							E2403	No Clinical Signs, Symptoms or Conditions	A051201	Device Dislodged or Dislocated	08/07/2022	08/07/2022
							F1901	Additional Surgery	A051201	Device Dislodged or Dislocated	08/07/2022	08/07/2022
7282	2022	All other reportable incidents	35852	Single-chamber implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem	25/08/2022	25/08/2022
									A070504	Premature Discharge of Battery	25/08/2022	25/08/2022
							F08	Hospitalization or Prolonged Hospitalization	A0705	Battery Problem	25/08/2022	25/08/2022
									A070504	Premature Discharge of Battery	25/08/2022	25/08/2022
							F1905	Device Revision or Replacement	A0705	Battery Problem	25/08/2022	25/08/2022
									A070504	Premature Discharge of Battery	25/08/2022	25/08/2022
7283	2022	All other reportable incidents	17882	Physiologic-monitoring defibrillation system	CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A1603	Failure of Device to Self-Test		
							F27	Problem Identified During Non-Clinical Procedure	A1603	Failure of Device to Self-Test		
7284	2022	All other reportable incidents	17882	Physiologic-monitoring defibrillation system	CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070803	Failure to Power Up		
							F27	Problem Identified During Non-Clinical Procedure	A070803	Failure to Power Up		
7285	2021	All other reportable incidents	35223	Endocardial/interventricular septal pacing lead	CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070909	Oversensing		
							F23	Unexpected Medical Intervention	A070909	Oversensing		
7286	2021	All other reportable incidents	17882	Physiologic-monitoring defibrillation system	CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A07	Electrical /Electronic Property Problem		
							F27	Problem Identified During Non-Clinical Procedure	A07	Electrical /Electronic Property Problem		
7287	2021	All other reportable incidents	17882	Physiologic-monitoring defibrillation system	CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A07	Electrical /Electronic Property Problem		
							F27	Problem Identified During Non-Clinical Procedure	A07	Electrical /Electronic Property Problem		
7288	2021	All other reportable incidents	47265	Dual-chamber implantable pacemaker, rate-responsive	CT141	Electrodes/Leads and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A0712	Pacing Problem	01/11/2021	01/11/2021
							F1903	Device Explantation	A0712	Pacing Problem	01/11/2021	01/11/2021
					CT145	Cardiovascular devices	E2403	No Clinical Signs, Symptoms or Conditions	A0712	Pacing Problem	01/11/2021	01/11/2021
							F1903	Device Explantation	A0712	Pacing Problem	01/11/2021	01/11/2021
					CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A0712	Pacing Problem	01/11/2021	01/11/2021
							F1903	Device Explantation	A0712	Pacing Problem	01/11/2021	01/11/2021
					CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A0712	Pacing Problem	01/11/2021	01/11/2021
							F1903	Device Explantation	A0712	Pacing Problem	01/11/2021	01/11/2021
					CT233	Surgical	E2403	No Clinical Signs, Symptoms or Conditions	A0712	Pacing Problem	01/11/2021	01/11/2021
							F1903	Device Explantation	A0712	Pacing Problem	01/11/2021	01/11/2021
					CT2406	Implantable	E2403	No Clinical Signs, Symptoms or Conditions	A0712	Pacing Problem	01/11/2021	01/11/2021
							F1903	Device Explantation	A0712	Pacing Problem	01/11/2021	01/11/2021
					CT2407	Active implantable	E2403	No Clinical Signs, Symptoms or Conditions	A0712	Pacing Problem	01/11/2021	01/11/2021
							F1903	Device Explantation	A0712	Pacing Problem	01/11/2021	01/11/2021
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A0712	Pacing Problem	01/11/2021	01/11/2021
							F1903	Device Explantation	A0712	Pacing Problem	01/11/2021	01/11/2021
					CT2727	Active implant devices/ancillaries	E2403	No Clinical Signs, Symptoms or Conditions	A0712	Pacing Problem	01/11/2021	01/11/2021
							F1903	Device Explantation	A0712	Pacing Problem	01/11/2021	01/11/2021
					CT2750	Implantable cardiac electrical stimulators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A0712	Pacing Problem	01/11/2021	01/11/2021
							F1903	Device Explantation	A0712	Pacing Problem	01/11/2021	01/11/2021
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A0712	Pacing Problem	01/11/2021	01/11/2021
							F1903	Device Explantation	A0712	Pacing Problem	01/11/2021	01/11/2021
					CT2985	Active	E2403	No Clinical Signs, Symptoms or Conditions	A0712	Pacing Problem	01/11/2021	01/11/2021
							F1903	Device Explantation	A0712	Pacing Problem	01/11/2021	01/11/2021
					CT3001	Implantable metal	E2403	No Clinical Signs, Symptoms or Conditions	A0712	Pacing Problem	01/11/2021	01/11/2021
							F1903	Device Explantation	A0712	Pacing Problem	01/11/2021	01/11/2021

					CT3002	Implantable synthetic polymer	E2403	No Clinical Signs, Symptoms or Conditions	A0712	Pacing Problem	01/11/2021	01/11/2021
							F1903	Device Explantation	A0712	Pacing Problem	01/11/2021	01/11/2021
					CT3006	Implantable inorganic material	E2403	No Clinical Signs, Symptoms or Conditions	A0712	Pacing Problem	01/11/2021	01/11/2021
							F1903	Device Explantation	A0712	Pacing Problem	01/11/2021	01/11/2021
					CT314	Electrically powered	E2403	No Clinical Signs, Symptoms or Conditions	A0712	Pacing Problem	01/11/2021	01/11/2021
							F1903	Device Explantation	A0712	Pacing Problem	01/11/2021	01/11/2021
					CT321	Long-term surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A0712	Pacing Problem	01/11/2021	01/11/2021
							F1903	Device Explantation	A0712	Pacing Problem	01/11/2021	01/11/2021
					CT449	Generators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A0712	Pacing Problem	01/11/2021	01/11/2021
							F1903	Device Explantation	A0712	Pacing Problem	01/11/2021	01/11/2021
					CT981	Single use	E2403	No Clinical Signs, Symptoms or Conditions	A0712	Pacing Problem	01/11/2021	01/11/2021
							F1903	Device Explantation	A0712	Pacing Problem	01/11/2021	01/11/2021
					CT983	Surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A0712	Pacing Problem	01/11/2021	01/11/2021
							F1903	Device Explantation	A0712	Pacing Problem	01/11/2021	01/11/2021
7289	2021	All other reportable incidents	47263	Cardiac resynchronization therapy implantable pacemaker	CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
7290	2021	All other reportable incidents	47263	Cardiac resynchronization therapy implantable pacemaker	CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A0908	Incorrect, Inadequate or Imprecise Result or Readings	02/11/2021	02/11/2021
							F08	Hospitalization or Prolonged Hospitalization	A0908	Incorrect, Inadequate or Imprecise Result or Readings	02/11/2021	02/11/2021
							F1903	Device Explantation	A0908	Incorrect, Inadequate or Imprecise Result or Readings	02/11/2021	02/11/2021
					CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A0908	Incorrect, Inadequate or Imprecise Result or Readings	02/11/2021	02/11/2021
							F08	Hospitalization or Prolonged Hospitalization	A0908	Incorrect, Inadequate or Imprecise Result or Readings	02/11/2021	02/11/2021
							F1903	Device Explantation	A0908	Incorrect, Inadequate or Imprecise Result or Readings	02/11/2021	02/11/2021
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A0908	Incorrect, Inadequate or Imprecise Result or Readings	02/11/2021	02/11/2021
							F08	Hospitalization or Prolonged Hospitalization	A0908	Incorrect, Inadequate or Imprecise Result or Readings	02/11/2021	02/11/2021
							F1903	Device Explantation	A0908	Incorrect, Inadequate or Imprecise Result or Readings	02/11/2021	02/11/2021
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A0908	Incorrect, Inadequate or Imprecise Result or Readings	02/11/2021	02/11/2021
							F08	Hospitalization or Prolonged Hospitalization	A0908	Incorrect, Inadequate or Imprecise Result or Readings	02/11/2021	02/11/2021
							F1903	Device Explantation	A0908	Incorrect, Inadequate or Imprecise Result or Readings	02/11/2021	02/11/2021
7291	2021	All other reportable incidents	35852	Single-chamber implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	20/12/2021	20/12/2021
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	20/12/2021	20/12/2021
							F1903	Device Explantation	A070504	Premature Discharge of Battery	20/12/2021	20/12/2021
7292	2021	All other reportable incidents	47265	Dual-chamber implantable pacemaker, rate-responsive	CT141	Electrodes/Leads and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	10/11/2021	10/11/2021
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	10/11/2021	10/11/2021
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	10/11/2021	10/11/2021
					CT145	Cardiovascular devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	10/11/2021	10/11/2021
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	10/11/2021	10/11/2021
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	10/11/2021	10/11/2021
					CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	10/11/2021	10/11/2021
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	10/11/2021	10/11/2021
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	10/11/2021	10/11/2021
					CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	10/11/2021	10/11/2021
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	10/11/2021	10/11/2021
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	10/11/2021	10/11/2021
					CT233	Surgical	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	10/11/2021	10/11/2021
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	10/11/2021	10/11/2021
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	10/11/2021	10/11/2021
					CT2406	Implantable	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	10/11/2021	10/11/2021
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	10/11/2021	10/11/2021
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	10/11/2021	10/11/2021
					CT2407	Active implantable	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	10/11/2021	10/11/2021
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	10/11/2021	10/11/2021
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	10/11/2021	10/11/2021
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	10/11/2021	10/11/2021
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	10/11/2021	10/11/2021
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	10/11/2021	10/11/2021
					CT2727	Active implant devices/ancillaries	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	10/11/2021	10/11/2021
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	10/11/2021	10/11/2021
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	10/11/2021	10/11/2021
					CT2750	Implantable cardiac electrical stimulators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	10/11/2021	10/11/2021
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	10/11/2021	10/11/2021
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	10/11/2021	10/11/2021
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	10/11/2021	10/11/2021
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	10/11/2021	10/11/2021
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	10/11/2021	10/11/2021
					CT2985	Active	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	10/11/2021	10/11/2021
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	10/11/2021	10/11/2021
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	10/11/2021	10/11/2021
					CT3001	Implantable metal	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	10/11/2021	10/11/2021
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	10/11/2021	10/11/2021
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	10/11/2021	10/11/2021
					CT3002	Implantable synthetic polymer	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	10/11/2021	10/11/2021
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	10/11/2021	10/11/2021
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	10/11/2021	10/11/2021

					CT3006	Implantable inorganic material	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	10/11/2021	10/11/2021
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	10/11/2021	10/11/2021
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	10/11/2021	10/11/2021
					CT314	Electrically powered	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	10/11/2021	10/11/2021
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	10/11/2021	10/11/2021
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	10/11/2021	10/11/2021
					CT321	Long-term surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	10/11/2021	10/11/2021
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	10/11/2021	10/11/2021
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	10/11/2021	10/11/2021
					CT449	Generators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	10/11/2021	10/11/2021
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	10/11/2021	10/11/2021
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	10/11/2021	10/11/2021
					CT981	Single use	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	10/11/2021	10/11/2021
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	10/11/2021	10/11/2021
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	10/11/2021	10/11/2021
					CT983	Surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	10/11/2021	10/11/2021
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	10/11/2021	10/11/2021
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	10/11/2021	10/11/2021
7293	2021	All other reportable incidents	47270	Cardiac resynchronization therapy implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	26/11/2021	26/11/2021
							F1901	Additional Surgery	A24	No Apparent Device Problem	26/11/2021	26/11/2021
7294	2021	All other reportable incidents	36241	Pacemaker/defibrillator lead	CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A072201	High Impedance		
							F26	No Health Consequences or Impact	A072201	High Impedance		
7295	2021	All other reportable incidents	35852	Single-chamber implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A072201	High Impedance		
							F26	No Health Consequences or Impact	A072201	High Impedance		
7296	2021	All other reportable incidents	17882	Physiologic-monitoring defibrillation system	CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A0719	Unexpected Shutdown		
							F26	No Health Consequences or Impact	A0719	Unexpected Shutdown		
7297	2022	All other reportable incidents	35852	Single-chamber implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
7298	2022	All other reportable incidents	47265	Dual-chamber implantable pacemaker, rate-responsive	CT141	Electrodes/Leads and associated devices	E2311	Discomfort	A070910	Undersensing		
									A071206	Pacing Asynchronously		
									A071302	Inappropriate/Inadequate Shock/Stimulation		
							E2403	No Clinical Signs, Symptoms or Conditions	A070910	Undersensing		
									A071206	Pacing Asynchronously		
									A071302	Inappropriate/Inadequate Shock/Stimulation		
							F26	No Health Consequences or Impact	A070910	Undersensing		
									A071206	Pacing Asynchronously		
									A071302	Inappropriate/Inadequate Shock/Stimulation		
					CT145	Cardiovascular devices	E2311	Discomfort	A070910	Undersensing		
									A071206	Pacing Asynchronously		
									A071302	Inappropriate/Inadequate Shock/Stimulation		
							E2403	No Clinical Signs, Symptoms or Conditions	A070910	Undersensing		
									A071206	Pacing Asynchronously		
									A071302	Inappropriate/Inadequate Shock/Stimulation		
							F26	No Health Consequences or Impact	A070910	Undersensing		
									A071206	Pacing Asynchronously		
									A071302	Inappropriate/Inadequate Shock/Stimulation		
					CT1684	Cardiac pacemakers and associated devices	E2311	Discomfort	A070910	Undersensing		
									A071206	Pacing Asynchronously		
									A071302	Inappropriate/Inadequate Shock/Stimulation		
							E2403	No Clinical Signs, Symptoms or Conditions	A070910	Undersensing		
									A071206	Pacing Asynchronously		
									A071302	Inappropriate/Inadequate Shock/Stimulation		
							F26	No Health Consequences or Impact	A070910	Undersensing		
									A071206	Pacing Asynchronously		
									A071302	Inappropriate/Inadequate Shock/Stimulation		
					CT1685	Cardiac pacemakers	E2311	Discomfort	A070910	Undersensing		
									A071206	Pacing Asynchronously		
									A071302	Inappropriate/Inadequate Shock/Stimulation		
							E2403	No Clinical Signs, Symptoms or Conditions	A070910	Undersensing		
									A071206	Pacing Asynchronously		
									A071302	Inappropriate/Inadequate Shock/Stimulation		
							F26	No Health Consequences or Impact	A070910	Undersensing		
									A071206	Pacing Asynchronously		
									A071302	Inappropriate/Inadequate Shock/Stimulation		
					CT233	Surgical	E2311	Discomfort	A070910	Undersensing		
									A071206	Pacing Asynchronously		
									A071302	Inappropriate/Inadequate Shock/Stimulation		
							E2403	No Clinical Signs, Symptoms or Conditions	A070910	Undersensing		
									A071206	Pacing Asynchronously		
									A071302	Inappropriate/Inadequate Shock/Stimulation		
							F26	No Health Consequences or Impact	A070910	Undersensing		
									A071206	Pacing Asynchronously		
									A071302	Inappropriate/Inadequate Shock/Stimulation		
					CT2406	Implantable	E2311	Discomfort	A070910	Undersensing		
									A071206	Pacing Asynchronously		
									A071302	Inappropriate/Inadequate Shock/Stimulation		
							E2403	No Clinical Signs, Symptoms or Conditions	A070910	Undersensing		
									A071206	Pacing Asynchronously		
									A071302	Inappropriate/Inadequate Shock/Stimulation		
							F26	No Health Consequences or Impact	A070910	Undersensing		
									A071206	Pacing Asynchronously		
									A071302	Inappropriate/Inadequate Shock/Stimulation		
					CT2407	Active implantable	E2311	Discomfort	A070910	Undersensing		
									A071206	Pacing Asynchronously		
									A071302	Inappropriate/Inadequate Shock/Stimulation		
							E2403	No Clinical Signs, Symptoms or Conditions	A070910	Undersensing		
									A071206	Pacing Asynchronously		
									A071302	Inappropriate/Inadequate Shock/Stimulation		

						F26	No Health Consequences or Impact	A070910	Undersensing		
								A071206	Pacing Asynchronously		
								A071302	Inappropriate/Inadequate Shock/Stimulation		
					CT257	Implantable cardiac pacemakers	E2311	Discomfort	A070910	Undersensing		
									A071206	Pacing Asynchronously		
									A071302	Inappropriate/Inadequate Shock/Stimulation		
							E2403	No Clinical Signs, Symptoms or Conditions	A070910	Undersensing		
									A071206	Pacing Asynchronously		
									A071302	Inappropriate/Inadequate Shock/Stimulation		
							F26	No Health Consequences or Impact	A070910	Undersensing		
									A071206	Pacing Asynchronously		
									A071302	Inappropriate/Inadequate Shock/Stimulation		
					CT2727	Active implant devices/ancillaries	E2311	Discomfort	A070910	Undersensing		
									A071206	Pacing Asynchronously		
									A071302	Inappropriate/Inadequate Shock/Stimulation		
							E2403	No Clinical Signs, Symptoms or Conditions	A070910	Undersensing		
									A071206	Pacing Asynchronously		
									A071302	Inappropriate/Inadequate Shock/Stimulation		
							F26	No Health Consequences or Impact	A070910	Undersensing		
									A071206	Pacing Asynchronously		
									A071302	Inappropriate/Inadequate Shock/Stimulation		
					CT2750	Implantable cardiac electrical stimulators and associated devices	E2311	Discomfort	A070910	Undersensing		
									A071206	Pacing Asynchronously		
									A071302	Inappropriate/Inadequate Shock/Stimulation		
							E2403	No Clinical Signs, Symptoms or Conditions	A070910	Undersensing		
									A071206	Pacing Asynchronously		
									A071302	Inappropriate/Inadequate Shock/Stimulation		
							F26	No Health Consequences or Impact	A070910	Undersensing		
									A071206	Pacing Asynchronously		
									A071302	Inappropriate/Inadequate Shock/Stimulation		
					CT2866	Implantable cardiac electrical stimulators	E2311	Discomfort	A070910	Undersensing		
									A071206	Pacing Asynchronously		
									A071302	Inappropriate/Inadequate Shock/Stimulation		
							E2403	No Clinical Signs, Symptoms or Conditions	A070910	Undersensing		
									A071206	Pacing Asynchronously		
									A071302	Inappropriate/Inadequate Shock/Stimulation		
							F26	No Health Consequences or Impact	A070910	Undersensing		
									A071206	Pacing Asynchronously		
									A071302	Inappropriate/Inadequate Shock/Stimulation		
					CT2985	Active	E2311	Discomfort	A070910	Undersensing		
									A071206	Pacing Asynchronously		
									A071302	Inappropriate/Inadequate Shock/Stimulation		
							E2403	No Clinical Signs, Symptoms or Conditions	A070910	Undersensing		
									A071206	Pacing Asynchronously		
									A071302	Inappropriate/Inadequate Shock/Stimulation		
							F26	No Health Consequences or Impact	A070910	Undersensing		
									A071206	Pacing Asynchronously		
									A071302	Inappropriate/Inadequate Shock/Stimulation		
					CT3001	Implantable metal	E2311	Discomfort	A070910	Undersensing		
									A071206	Pacing Asynchronously		
									A071302	Inappropriate/Inadequate Shock/Stimulation		
							E2403	No Clinical Signs, Symptoms or Conditions	A070910	Undersensing		
									A071206	Pacing Asynchronously		
									A071302	Inappropriate/Inadequate Shock/Stimulation		
							F26	No Health Consequences or Impact	A070910	Undersensing		
									A071206	Pacing Asynchronously		
									A071302	Inappropriate/Inadequate Shock/Stimulation		
					CT3002	Implantable synthetic polymer	E2311	Discomfort	A070910	Undersensing		
									A071206	Pacing Asynchronously		
									A071302	Inappropriate/Inadequate Shock/Stimulation		
							E2403	No Clinical Signs, Symptoms or Conditions	A070910	Undersensing		
									A071206	Pacing Asynchronously		
									A071302	Inappropriate/Inadequate Shock/Stimulation		
							F26	No Health Consequences or Impact	A070910	Undersensing		
									A071206	Pacing Asynchronously		
									A071302	Inappropriate/Inadequate Shock/Stimulation		
					CT3006	Implantable inorganic material	E2311	Discomfort	A070910	Undersensing		
									A071206	Pacing Asynchronously		
									A071302	Inappropriate/Inadequate Shock/Stimulation		
							E2403	No Clinical Signs, Symptoms or Conditions	A070910	Undersensing		
									A071206	Pacing Asynchronously		
									A071302	Inappropriate/Inadequate Shock/Stimulation		
							F26	No Health Consequences or Impact	A070910	Undersensing		
									A071206	Pacing Asynchronously		
									A071302	Inappropriate/Inadequate Shock/Stimulation		
					CT314	Electrically powered	E2311	Discomfort	A070910	Undersensing		
									A071206	Pacing Asynchronously		
									A071302	Inappropriate/Inadequate Shock/Stimulation		
							E2403	No Clinical Signs, Symptoms or Conditions	A070910	Undersensing		
									A071206	Pacing Asynchronously		
									A071302	Inappropriate/Inadequate Shock/Stimulation		
							F26	No Health Consequences or Impact	A070910	Undersensing		
									A071206	Pacing Asynchronously		
									A071302	Inappropriate/Inadequate Shock/Stimulation		
					CT321	Long-term surgical invasive	E2311	Discomfort	A070910	Undersensing		
									A071206	Pacing Asynchronously		
									A071302	Inappropriate/Inadequate Shock/Stimulation		
							E2403	No Clinical Signs, Symptoms or Conditions	A070910	Undersensing		

								A071206	Pacing Asynchronously		
								A071302	Inappropriate/Inadequate Shock/Stimulation		
						F26	No Health Consequences or Impact	A070910	Undersensing		
								A071206	Pacing Asynchronously		
								A071302	Inappropriate/Inadequate Shock/Stimulation		
					CT449	Generators and associated devices	E2311	Discomfort	A070910	Undersensing		
									A071206	Pacing Asynchronously		
									A071302	Inappropriate/Inadequate Shock/Stimulation		
							E2403	No Clinical Signs, Symptoms or Conditions	A070910	Undersensing		
									A071206	Pacing Asynchronously		
									A071302	Inappropriate/Inadequate Shock/Stimulation		
							F26	No Health Consequences or Impact	A070910	Undersensing		
									A071206	Pacing Asynchronously		
									A071302	Inappropriate/Inadequate Shock/Stimulation		
					CT981	Single use	E2311	Discomfort	A070910	Undersensing		
									A071206	Pacing Asynchronously		
									A071302	Inappropriate/Inadequate Shock/Stimulation		
							E2403	No Clinical Signs, Symptoms or Conditions	A070910	Undersensing		
									A071206	Pacing Asynchronously		
									A071302	Inappropriate/Inadequate Shock/Stimulation		
							F26	No Health Consequences or Impact	A070910	Undersensing		
									A071206	Pacing Asynchronously		
									A071302	Inappropriate/Inadequate Shock/Stimulation		
					CT983	Surgical invasive	E2311	Discomfort	A070910	Undersensing		
									A071206	Pacing Asynchronously		
									A071302	Inappropriate/Inadequate Shock/Stimulation		
							E2403	No Clinical Signs, Symptoms or Conditions	A070910	Undersensing		
									A071206	Pacing Asynchronously		
									A071302	Inappropriate/Inadequate Shock/Stimulation		
							F26	No Health Consequences or Impact	A070910	Undersensing		
									A071206	Pacing Asynchronously		
									A071302	Inappropriate/Inadequate Shock/Stimulation		
7299	2021	All other reportable incidents	35852	Single-chamber implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem	02/02/2022	02/02/2022
									A070504	Premature Discharge of Battery	02/02/2022	02/02/2022
							F08	Hospitalization or Prolonged Hospitalization	A0705	Battery Problem	02/02/2022	02/02/2022
									A070504	Premature Discharge of Battery	02/02/2022	02/02/2022
							F1905	Device Revision or Replacement	A0705	Battery Problem	02/02/2022	02/02/2022
									A070504	Premature Discharge of Battery	02/02/2022	02/02/2022
7300	2021	All other reportable incidents	47263	Cardiac resynchronization therapy implantable pacemaker	CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070910	Undersensing		
							F15	Recognized Device or Procedural Complication	A070910	Undersensing		
					CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A070910	Undersensing		
							F15	Recognized Device or Procedural Complication	A070910	Undersensing		
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A070910	Undersensing		
							F15	Recognized Device or Procedural Complication	A070910	Undersensing		
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A070910	Undersensing		
							F15	Recognized Device or Procedural Complication	A070910	Undersensing		
7301	2021	All other reportable incidents	47267	Single-chamber implantable pacemaker, rate-responsive	CT141	Electrodes/Leads and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	16/12/2021	16/12/2021
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	16/12/2021	16/12/2021
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	16/12/2021	16/12/2021
					CT145	Cardiovascular devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	16/12/2021	16/12/2021
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	16/12/2021	16/12/2021
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	16/12/2021	16/12/2021
					CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	16/12/2021	16/12/2021
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	16/12/2021	16/12/2021
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	16/12/2021	16/12/2021
					CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	16/12/2021	16/12/2021
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	16/12/2021	16/12/2021
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	16/12/2021	16/12/2021
					CT233	Surgical	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	16/12/2021	16/12/2021
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	16/12/2021	16/12/2021
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	16/12/2021	16/12/2021
					CT2406	Implantable	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	16/12/2021	16/12/2021
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	16/12/2021	16/12/2021
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	16/12/2021	16/12/2021
					CT2407	Active implantable	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	16/12/2021	16/12/2021
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	16/12/2021	16/12/2021
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	16/12/2021	16/12/2021
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	16/12/2021	16/12/2021
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	16/12/2021	16/12/2021
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	16/12/2021	16/12/2021
					CT2727	Active implant devices/ancillaries	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	16/12/2021	16/12/2021
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	16/12/2021	16/12/2021
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	16/12/2021	16/12/2021
					CT2750	Implantable cardiac electrical stimulators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	16/12/2021	16/12/2021
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	16/12/2021	16/12/2021
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	16/12/2021	16/12/2021
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	16/12/2021	16/12/2021
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	16/12/2021	16/12/2021
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	16/12/2021	16/12/2021
					CT2985	Active	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	16/12/2021	16/12/2021
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	16/12/2021	16/12/2021
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	16/12/2021	16/12/2021
					CT3001	Implantable metal	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	16/12/2021	16/12/2021
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	16/12/2021	16/12/2021
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	16/12/2021	16/12/2021
					CT3002	Implantable synthetic polymer	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	16/12/2021	16/12/2021
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	16/12/2021	16/12/2021
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	16/12/2021	16/12/2021

					CT3006	Implantable inorganic material	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	16/12/2021	16/12/2021
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	16/12/2021	16/12/2021
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	16/12/2021	16/12/2021
					CT314	Electrically powered	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	16/12/2021	16/12/2021
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	16/12/2021	16/12/2021
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	16/12/2021	16/12/2021
					CT321	Long-term surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	16/12/2021	16/12/2021
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	16/12/2021	16/12/2021
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	16/12/2021	16/12/2021
					CT449	Generators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	16/12/2021	16/12/2021
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	16/12/2021	16/12/2021
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	16/12/2021	16/12/2021
					CT981	Single use	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	16/12/2021	16/12/2021
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	16/12/2021	16/12/2021
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	16/12/2021	16/12/2021
					CT983	Surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	16/12/2021	16/12/2021
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	16/12/2021	16/12/2021
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	16/12/2021	16/12/2021
7302	2022	All other reportable incidents	17882	Physiologic-monitoring defibrillation system	CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070803	Failure to Power Up		
							F27	Problem Identified During Non-Clinical Procedure	A070803	Failure to Power Up		
7303	2021	All other reportable incidents	36241	Pacemaker/defibrillator lead	CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A04	Material Integrity Problem		
									A070909	Oversensing		
									A070910	Undersensing		
									A072202	Low Impedance		
									A090801	Signal Artifact/Noise		
							F08	Hospitalization or Prolonged Hospitalization	A04	Material Integrity Problem		
									A070909	Oversensing		
									A070910	Undersensing		
									A072202	Low Impedance		
									A090801	Signal Artifact/Noise		
7304	2021	All other reportable incidents	35223	Endocardial/interventricular septal pacing lead	CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070909	Oversensing	03/12/2021	03/12/2021
									A090801	Signal Artifact/Noise	03/12/2021	03/12/2021
							F26	No Health Consequences or Impact	A070909	Oversensing	03/12/2021	03/12/2021
									A090801	Signal Artifact/Noise	03/12/2021	03/12/2021
7305	2021	All other reportable incidents	35223	Endocardial/interventricular septal pacing lead	CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A05	Mechanical Problem		
									A070909	Oversensing		
									A070910	Undersensing		
									A072202	Low Impedance		
									A090801	Signal Artifact/Noise		
							F26	No Health Consequences or Impact	A05	Mechanical Problem		
									A070909	Oversensing		
									A070910	Undersensing		
									A072202	Low Impedance		
									A090801	Signal Artifact/Noise		
7306	2021	All other reportable incidents	37265	Dual-chamber implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A070910	Undersensing		
									A1301	Failure to Read Input Signal		
							F08	Hospitalization or Prolonged Hospitalization	A070910	Undersensing		
									A1301	Failure to Read Input Signal		
							F1905	Device Revision or Replacement	A070910	Undersensing		
									A1301	Failure to Read Input Signal		
7307	2021	All other reportable incidents	47265	Dual-chamber implantable pacemaker, rate-responsive	CT141	Electrodes/Leads and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	09/03/2022	09/03/2022
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	09/03/2022	09/03/2022
							F1903	Device Explantation	A070504	Premature Discharge of Battery	09/03/2022	09/03/2022
					CT145	Cardiovascular devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	09/03/2022	09/03/2022
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	09/03/2022	09/03/2022
							F1903	Device Explantation	A070504	Premature Discharge of Battery	09/03/2022	09/03/2022
					CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	09/03/2022	09/03/2022
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	09/03/2022	09/03/2022
							F1903	Device Explantation	A070504	Premature Discharge of Battery	09/03/2022	09/03/2022
					CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	09/03/2022	09/03/2022
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	09/03/2022	09/03/2022
							F1903	Device Explantation	A070504	Premature Discharge of Battery	09/03/2022	09/03/2022
					CT233	Surgical	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	09/03/2022	09/03/2022
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	09/03/2022	09/03/2022
							F1903	Device Explantation	A070504	Premature Discharge of Battery	09/03/2022	09/03/2022
					CT2406	Implantable	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	09/03/2022	09/03/2022
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	09/03/2022	09/03/2022
							F1903	Device Explantation	A070504	Premature Discharge of Battery	09/03/2022	09/03/2022
					CT2407	Active implantable	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	09/03/2022	09/03/2022
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	09/03/2022	09/03/2022
							F1903	Device Explantation	A070504	Premature Discharge of Battery	09/03/2022	09/03/2022
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	09/03/2022	09/03/2022
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	09/03/2022	09/03/2022
							F1903	Device Explantation	A070504	Premature Discharge of Battery	09/03/2022	09/03/2022
					CT2727	Active implant devices/ancillaries	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	09/03/2022	09/03/2022
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	09/03/2022	09/03/2022
							F1903	Device Explantation	A070504	Premature Discharge of Battery	09/03/2022	09/03/2022
					CT2750	Implantable cardiac electrical stimulators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	09/03/2022	09/03/2022
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	09/03/2022	09/03/2022
							F1903	Device Explantation	A070504	Premature Discharge of Battery	09/03/2022	09/03/2022
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	09/03/2022	09/03/2022
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	09/03/2022	09/03/2022
							F1903	Device Explantation	A070504	Premature Discharge of Battery	09/03/2022	09/03/2022
					CT2985	Active	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	09/03/2022	09/03/2022
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	09/03/2022	09/03/2022
							F1903	Device Explantation	A070504	Premature Discharge of Battery	09/03/2022	09/03/2022
					CT3001	Implantable metal	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	09/03/2022	09/03/2022
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	09/03/2022	09/03/2022

							F1903	Device Explantation	A070504	Premature Discharge of Battery	09/03/2022	09/03/2022
					CT3002	Implantable synthetic polymer	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	09/03/2022	09/03/2022
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	09/03/2022	09/03/2022
							F1903	Device Explantation	A070504	Premature Discharge of Battery	09/03/2022	09/03/2022
					CT3006	Implantable inorganic material	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	09/03/2022	09/03/2022
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	09/03/2022	09/03/2022
							F1903	Device Explantation	A070504	Premature Discharge of Battery	09/03/2022	09/03/2022
					CT314	Electrically powered	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	09/03/2022	09/03/2022
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	09/03/2022	09/03/2022
							F1903	Device Explantation	A070504	Premature Discharge of Battery	09/03/2022	09/03/2022
					CT321	Long-term surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	09/03/2022	09/03/2022
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	09/03/2022	09/03/2022
							F1903	Device Explantation	A070504	Premature Discharge of Battery	09/03/2022	09/03/2022
					CT449	Generators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	09/03/2022	09/03/2022
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	09/03/2022	09/03/2022
							F1903	Device Explantation	A070504	Premature Discharge of Battery	09/03/2022	09/03/2022
					CT981	Single use	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	09/03/2022	09/03/2022
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	09/03/2022	09/03/2022
							F1903	Device Explantation	A070504	Premature Discharge of Battery	09/03/2022	09/03/2022
					CT983	Surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	09/03/2022	09/03/2022
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	09/03/2022	09/03/2022
							F1903	Device Explantation	A070504	Premature Discharge of Battery	09/03/2022	09/03/2022
7308	2021	All other reportable incidents	35223	Endocardial/interventricular septal pacing lead	CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A0401	Break		
									A070101	Failure to Capture		
									A072202	Low Impedance		
							F1901	Additional Surgery	A0401	Break		
									A070101	Failure to Capture		
									A072202	Low Impedance		
7309	2021	All other reportable incidents	35223	Endocardial/interventricular septal pacing lead	CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070909	Oversensing		
							F1901	Additional Surgery	A070909	Oversensing		
7310	2021	All other reportable incidents	47270	Cardiac resynchronization therapy implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A070603	Failure to Charge	04/01/2022	04/01/2022
							F1901	Additional Surgery	A070603	Failure to Charge	04/01/2022	04/01/2022
7311	2021	All other reportable incidents	35223	Endocardial/interventricular septal pacing lead	CT1684	Cardiac pacemakers and associated devices	E0112	Dizziness	A070909	Oversensing		
									A23	Use of Device Problem		
							E2403	No Clinical Signs, Symptoms or Conditions	A070909	Oversensing		
									A23	Use of Device Problem		
							F1901	Additional Surgery	A070909	Oversensing		
									A23	Use of Device Problem		
7312	2022	All other reportable incidents	35852	Single-chamber implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem	29/09/2022	29/09/2022
									A070504	Premature Discharge of Battery	29/09/2022	29/09/2022
							F08	Hospitalization or Prolonged Hospitalization	A0705	Battery Problem	29/09/2022	29/09/2022
									A070504	Premature Discharge of Battery	29/09/2022	29/09/2022
							F1905	Device Revision or Replacement	A0705	Battery Problem	29/09/2022	29/09/2022
									A070504	Premature Discharge of Battery	29/09/2022	29/09/2022
7313	2022	All other reportable incidents	47267	Single-chamber implantable pacemaker, rate-responsive	CT141	Electrodes/Leads and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT145	Cardiovascular devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT233	Surgical	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT2406	Implantable	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT2407	Active implantable	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT2727	Active implant devices/ancillaries	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT2750	Implantable cardiac electrical stimulators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT2985	Active	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT3001	Implantable metal	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT3002	Implantable synthetic polymer	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT3006	Implantable inorganic material	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT314	Electrically powered	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT321	Long-term surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT449	Generators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT981	Single use	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT983	Surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
7314	2022	All other reportable incidents	47263	Cardiac resynchronization therapy implantable pacemaker	CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A0908	Incorrect, Inadequate or Imprecise Result or Readings	21/07/2022	21/07/2022
							F08	Hospitalization or Prolonged Hospitalization	A0908	Incorrect, Inadequate or Imprecise Result or Readings	21/07/2022	21/07/2022

							F1905	Device Revision or Replacement	A0908	Incorrect, Inadequate or Imprecise Result or Readings	21/07/2022	21/07/2022
					CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A0908	Incorrect, Inadequate or Imprecise Result or Readings	21/07/2022	21/07/2022
							F08	Hospitalization or Prolonged Hospitalization	A0908	Incorrect, Inadequate or Imprecise Result or Readings	21/07/2022	21/07/2022
							F1905	Device Revision or Replacement	A0908	Incorrect, Inadequate or Imprecise Result or Readings	21/07/2022	21/07/2022
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A0908	Incorrect, Inadequate or Imprecise Result or Readings	21/07/2022	21/07/2022
							F08	Hospitalization or Prolonged Hospitalization	A0908	Incorrect, Inadequate or Imprecise Result or Readings	21/07/2022	21/07/2022
							F1905	Device Revision or Replacement	A0908	Incorrect, Inadequate or Imprecise Result or Readings	21/07/2022	21/07/2022
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A0908	Incorrect, Inadequate or Imprecise Result or Readings	21/07/2022	21/07/2022
							F08	Hospitalization or Prolonged Hospitalization	A0908	Incorrect, Inadequate or Imprecise Result or Readings	21/07/2022	21/07/2022
							F1905	Device Revision or Replacement	A0908	Incorrect, Inadequate or Imprecise Result or Readings	21/07/2022	21/07/2022
7315	2021	All other reportable incidents	37265	Dual-chamber implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A0908	Incorrect, Inadequate or Imprecise Result or Readings	01/12/2021	01/12/2021
							F08	Hospitalization or Prolonged Hospitalization	A0908	Incorrect, Inadequate or Imprecise Result or Readings	01/12/2021	01/12/2021
							F1905	Device Revision or Replacement	A0908	Incorrect, Inadequate or Imprecise Result or Readings	01/12/2021	01/12/2021
7316	2022	All other reportable incidents	35852	Single-chamber implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
7317	2021	All other reportable incidents	17882	Physiologic-monitoring defibrillation system	CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070908	Failure to Sense		
							F26	No Health Consequences or Impact	A070908	Failure to Sense		
7318	2022	All other reportable incidents	17882	Physiologic-monitoring defibrillation system	CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A07	Electrical /Electronic Property Problem		
							F27	Problem Identified During Non-Clinical Procedure	A07	Electrical /Electronic Property Problem		
7319	2021	All other reportable incidents	17882	Physiologic-monitoring defibrillation system	CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A0719	Unexpected Shutdown		
							F26	No Health Consequences or Impact	A0719	Unexpected Shutdown		
7320	2022	All other reportable incidents	35223	Endocardial/interventricular septal pacing lead	CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070910	Undersensing		
							F26	No Health Consequences or Impact	A070910	Undersensing		
7321	2022	All other reportable incidents	17882	Physiologic-monitoring defibrillation system	CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070603	Failure to Charge		
							F27	Problem Identified During Non-Clinical Procedure	A070603	Failure to Charge		
7322	2021	All other reportable incidents	35223	Endocardial/interventricular septal pacing lead	CT1684	Cardiac pacemakers and associated devices	E060104	Bradycardia	A070101	Failure to Capture	07/12/2021	
									A070102	High Capture Threshold	07/12/2021	
									A070909	Oversensing	07/12/2021	
									A072201	High Impedance	07/12/2021	
									A090801	Signal Artifact/Noise	07/12/2021	
							E2403	No Clinical Signs, Symptoms or Conditions	A070101	Failure to Capture	07/12/2021	
									A070102	High Capture Threshold	07/12/2021	
									A070909	Oversensing	07/12/2021	
									A072201	High Impedance	07/12/2021	
									A090801	Signal Artifact/Noise	07/12/2021	
							F11	Non-Serious Injury/ Illness/ Impairment	A070101	Failure to Capture	07/12/2021	
									A070102	High Capture Threshold	07/12/2021	
									A070909	Oversensing	07/12/2021	
									A072201	High Impedance	07/12/2021	
									A090801	Signal Artifact/Noise	07/12/2021	
							F19	Surgical Intervention	A070101	Failure to Capture	07/12/2021	
									A070102	High Capture Threshold	07/12/2021	
									A070909	Oversensing	07/12/2021	
									A072201	High Impedance	07/12/2021	
									A090801	Signal Artifact/Noise	07/12/2021	
7323	2022	All other reportable incidents	47263	Cardiac resynchronization therapy implantable pacemaker	CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A0908	Incorrect, Inadequate or Imprecise Result or Readings	21/07/2022	21/07/2022
							F08	Hospitalization or Prolonged Hospitalization	A0908	Incorrect, Inadequate or Imprecise Result or Readings	21/07/2022	21/07/2022
							F1905	Device Revision or Replacement	A0908	Incorrect, Inadequate or Imprecise Result or Readings	21/07/2022	21/07/2022
					CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A0908	Incorrect, Inadequate or Imprecise Result or Readings	21/07/2022	21/07/2022
							F08	Hospitalization or Prolonged Hospitalization	A0908	Incorrect, Inadequate or Imprecise Result or Readings	21/07/2022	21/07/2022
							F1905	Device Revision or Replacement	A0908	Incorrect, Inadequate or Imprecise Result or Readings	21/07/2022	21/07/2022
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A0908	Incorrect, Inadequate or Imprecise Result or Readings	21/07/2022	21/07/2022
							F08	Hospitalization or Prolonged Hospitalization	A0908	Incorrect, Inadequate or Imprecise Result or Readings	21/07/2022	21/07/2022
							F1905	Device Revision or Replacement	A0908	Incorrect, Inadequate or Imprecise Result or Readings	21/07/2022	21/07/2022
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A0908	Incorrect, Inadequate or Imprecise Result or Readings	21/07/2022	21/07/2022
							F08	Hospitalization or Prolonged Hospitalization	A0908	Incorrect, Inadequate or Imprecise Result or Readings	21/07/2022	21/07/2022
							F1905	Device Revision or Replacement	A0908	Incorrect, Inadequate or Imprecise Result or Readings	21/07/2022	21/07/2022
7324	2022	All other reportable incidents	47263	Cardiac resynchronization therapy implantable pacemaker	CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem	04/11/2022	04/11/2022
									A070504	Premature Discharge of Battery	04/11/2022	04/11/2022
							F08	Hospitalization or Prolonged Hospitalization	A0705	Battery Problem	04/11/2022	04/11/2022
									A070504	Premature Discharge of Battery	04/11/2022	04/11/2022
							F1905	Device Revision or Replacement	A0705	Battery Problem	04/11/2022	04/11/2022
									A070504	Premature Discharge of Battery	04/11/2022	04/11/2022
					CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem	04/11/2022	04/11/2022
									A070504	Premature Discharge of Battery	04/11/2022	04/11/2022
							F08	Hospitalization or Prolonged Hospitalization	A0705	Battery Problem	04/11/2022	04/11/2022
									A070504	Premature Discharge of Battery	04/11/2022	04/11/2022

							F1905	Device Revision or Replacement	A0705	Battery Problem	04/11/2022	04/11/2022
									A070504	Premature Discharge of Battery	04/11/2022	04/11/2022
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem	04/11/2022	04/11/2022
									A070504	Premature Discharge of Battery	04/11/2022	04/11/2022
							F08	Hospitalization or Prolonged Hospitalization	A0705	Battery Problem	04/11/2022	04/11/2022
									A070504	Premature Discharge of Battery	04/11/2022	04/11/2022
							F1905	Device Revision or Replacement	A0705	Battery Problem	04/11/2022	04/11/2022
									A070504	Premature Discharge of Battery	04/11/2022	04/11/2022
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem	04/11/2022	04/11/2022
									A070504	Premature Discharge of Battery	04/11/2022	04/11/2022
							F08	Hospitalization or Prolonged Hospitalization	A0705	Battery Problem	04/11/2022	04/11/2022
									A070504	Premature Discharge of Battery	04/11/2022	04/11/2022
							F1905	Device Revision or Replacement	A0705	Battery Problem	04/11/2022	04/11/2022
									A070504	Premature Discharge of Battery	04/11/2022	04/11/2022
7325	2022	All other reportable incidents	35223	Endocardial/interventricular septal pacing lead	CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070909	Oversensing		
							F26	No Health Consequences or Impact	A070909	Oversensing		
7326	2022	All other reportable incidents	35223	Endocardial/interventricular septal pacing lead	CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A0401	Break		
									A070901	Decreased Sensitivity		
									A070909	Oversensing		
									A070910	Undersensing		
									A0722	Impedance Problem		
									A090801	Signal Artifact/Noise		
							F26	No Health Consequences or Impact	A0401	Break		
									A070901	Decreased Sensitivity		
									A070909	Oversensing		
									A070910	Undersensing		
									A0722	Impedance Problem		
									A090801	Signal Artifact/Noise		
7327	2022	All other reportable incidents	35223	Endocardial/interventricular septal pacing lead	CT1684	Cardiac pacemakers and associated devices	E0112	Dizziness	A070102	High Capture Threshold	28/07/2022	28/07/2022
									A0722	Impedance Problem	28/07/2022	28/07/2022
							E011903	Syncope/Fainting	A070102	High Capture Threshold	28/07/2022	28/07/2022
									A0722	Impedance Problem	28/07/2022	28/07/2022
							E2403	No Clinical Signs, Symptoms or Conditions	A070102	High Capture Threshold	28/07/2022	28/07/2022
									A0722	Impedance Problem	28/07/2022	28/07/2022
							F1203	Life Threatening Illness or Injury	A070102	High Capture Threshold	28/07/2022	28/07/2022
									A0722	Impedance Problem	28/07/2022	28/07/2022
7328	2022	All other reportable incidents	35223	Endocardial/interventricular septal pacing lead	CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A0722	Impedance Problem		
							F26	No Health Consequences or Impact	A0722	Impedance Problem		
7329	2022	All other reportable incidents	35223	Endocardial/interventricular septal pacing lead	CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070102	High Capture Threshold		
									A0722	Impedance Problem		
							F26	No Health Consequences or Impact	A070102	High Capture Threshold		
									A0722	Impedance Problem		
7330	2022	All other reportable incidents	35223	Endocardial/interventricular septal pacing lead	CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A072201	High Impedance		
							F26	No Health Consequences or Impact	A072201	High Impedance		
7331	2022	All other reportable incidents	35223	Endocardial/interventricular septal pacing lead	CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070909	Oversensing		
									A072001	Electromagnetic Interference		
									A090801	Signal Artifact/Noise		
							F26	No Health Consequences or Impact	A070909	Oversensing		
									A072001	Electromagnetic Interference		
									A090801	Signal Artifact/Noise		
7332	2021	All other reportable incidents	35852	Single-chamber implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A070909	Oversensing		
							F23	Unexpected Medical Intervention	A070909	Oversensing		
7333	2021	All other reportable incidents	37265	Dual-chamber implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A070909	Oversensing		
							F15	Recognized Device or Procedural Complication	A070909	Oversensing		
7334	2022	All other reportable incidents	35852	Single-chamber implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A072201	High Impedance		
							F26	No Health Consequences or Impact	A072201	High Impedance		
7335	2021	All other reportable incidents	36241	Pacemaker/defibrillator lead	CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070909	Oversensing		
							F15	Recognized Device or Procedural Complication	A070909	Oversensing		
7336	2022	All other reportable incidents	36241	Pacemaker/defibrillator lead	CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A072201	High Impedance		
							F26	No Health Consequences or Impact	A072201	High Impedance		
7337	2021	All other reportable incidents	17882	Physiologic-monitoring defibrillation system	CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070803	Failure to Power Up		
							F27	Problem Identified During Non-Clinical Procedure	A070803	Failure to Power Up		
7338	2021	All other reportable incidents	47267	Single-chamber implantable pacemaker, rate-responsive	CT141	Electrodes/Leads and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT145	Cardiovascular devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT233	Surgical	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT2406	Implantable	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT2407	Active implantable	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT2727	Active implant devices/ancillaries	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT2750	Implantable cardiac electrical stimulators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT2985	Active	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT3001	Implantable metal	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		

					CT3002	Implantable synthetic polymer	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT3006	Implantable inorganic material	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT314	Electrically powered	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT321	Long-term surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT449	Generators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT981	Single use	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT983	Surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
7339	2021	All other reportable incidents	35852	Single-chamber implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
7340	2021	All other reportable incidents	47267	Single-chamber implantable pacemaker, rate-responsive	CT141	Electrodes/Leads and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	30/11/2021	30/11/2021
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	30/11/2021	30/11/2021
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	30/11/2021	30/11/2021
					CT145	Cardiovascular devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	30/11/2021	30/11/2021
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	30/11/2021	30/11/2021
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	30/11/2021	30/11/2021
					CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	30/11/2021	30/11/2021
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	30/11/2021	30/11/2021
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	30/11/2021	30/11/2021
					CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	30/11/2021	30/11/2021
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	30/11/2021	30/11/2021
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	30/11/2021	30/11/2021
					CT233	Surgical	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	30/11/2021	30/11/2021
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	30/11/2021	30/11/2021
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	30/11/2021	30/11/2021
					CT2406	Implantable	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	30/11/2021	30/11/2021
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	30/11/2021	30/11/2021
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	30/11/2021	30/11/2021
					CT2407	Active implantable	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	30/11/2021	30/11/2021
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	30/11/2021	30/11/2021
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	30/11/2021	30/11/2021
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	30/11/2021	30/11/2021
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	30/11/2021	30/11/2021
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	30/11/2021	30/11/2021
					CT2727	Active implant devices/ancillaries	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	30/11/2021	30/11/2021
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	30/11/2021	30/11/2021
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	30/11/2021	30/11/2021
					CT2750	Implantable cardiac electrical stimulators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	30/11/2021	30/11/2021
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	30/11/2021	30/11/2021
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	30/11/2021	30/11/2021
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	30/11/2021	30/11/2021
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	30/11/2021	30/11/2021
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	30/11/2021	30/11/2021
					CT2985	Active	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	30/11/2021	30/11/2021
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	30/11/2021	30/11/2021
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	30/11/2021	30/11/2021
					CT3001	Implantable metal	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	30/11/2021	30/11/2021
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	30/11/2021	30/11/2021
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	30/11/2021	30/11/2021
					CT3002	Implantable synthetic polymer	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	30/11/2021	30/11/2021
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	30/11/2021	30/11/2021
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	30/11/2021	30/11/2021
					CT3006	Implantable inorganic material	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	30/11/2021	30/11/2021
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	30/11/2021	30/11/2021
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	30/11/2021	30/11/2021
					CT314	Electrically powered	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	30/11/2021	30/11/2021
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	30/11/2021	30/11/2021
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	30/11/2021	30/11/2021
					CT321	Long-term surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	30/11/2021	30/11/2021
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	30/11/2021	30/11/2021
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	30/11/2021	30/11/2021
					CT449	Generators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	30/11/2021	30/11/2021
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	30/11/2021	30/11/2021
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	30/11/2021	30/11/2021
					CT981	Single use	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	30/11/2021	30/11/2021
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	30/11/2021	30/11/2021
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	30/11/2021	30/11/2021
					CT983	Surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	30/11/2021	30/11/2021
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	30/11/2021	30/11/2021
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	30/11/2021	30/11/2021
7341	2022	All other reportable incidents	17882	Physiologic-monitoring defibrillation system	CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070803	Failure to Power Up		
							F27	Problem Identified During Non-Clinical Procedure	A070803	Failure to Power Up		
7342	2021	All other reportable incidents	47265	Dual-chamber implantable pacemaker, rate-responsive	CT141	Electrodes/Leads and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	30/11/2021	30/11/2021
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	30/11/2021	30/11/2021
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	30/11/2021	30/11/2021
					CT145	Cardiovascular devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	30/11/2021	30/11/2021
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	30/11/2021	30/11/2021
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	30/11/2021	30/11/2021
					CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	30/11/2021	30/11/2021
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	30/11/2021	30/11/2021
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	30/11/2021	30/11/2021
					CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	30/11/2021	30/11/2021

							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	30/11/2021	30/11/2021
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	30/11/2021	30/11/2021
					CT233	Surgical	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	30/11/2021	30/11/2021
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	30/11/2021	30/11/2021
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	30/11/2021	30/11/2021
					CT2406	Implantable	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	30/11/2021	30/11/2021
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	30/11/2021	30/11/2021
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	30/11/2021	30/11/2021
					CT2407	Active implantable	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	30/11/2021	30/11/2021
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	30/11/2021	30/11/2021
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	30/11/2021	30/11/2021
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	30/11/2021	30/11/2021
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	30/11/2021	30/11/2021
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	30/11/2021	30/11/2021
					CT2727	Active implant devices/ancillaries	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	30/11/2021	30/11/2021
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	30/11/2021	30/11/2021
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	30/11/2021	30/11/2021
					CT2750	Implantable cardiac electrical stimulators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	30/11/2021	30/11/2021
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	30/11/2021	30/11/2021
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	30/11/2021	30/11/2021
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	30/11/2021	30/11/2021
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	30/11/2021	30/11/2021
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	30/11/2021	30/11/2021
					CT2985	Active	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	30/11/2021	30/11/2021
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	30/11/2021	30/11/2021
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	30/11/2021	30/11/2021
					CT3001	Implantable metal	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	30/11/2021	30/11/2021
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	30/11/2021	30/11/2021
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	30/11/2021	30/11/2021
					CT3002	Implantable synthetic polymer	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	30/11/2021	30/11/2021
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	30/11/2021	30/11/2021
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	30/11/2021	30/11/2021
					CT3006	Implantable inorganic material	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	30/11/2021	30/11/2021
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	30/11/2021	30/11/2021
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	30/11/2021	30/11/2021
					CT314	Electrically powered	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	30/11/2021	30/11/2021
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	30/11/2021	30/11/2021
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	30/11/2021	30/11/2021
					CT321	Long-term surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	30/11/2021	30/11/2021
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	30/11/2021	30/11/2021
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	30/11/2021	30/11/2021
					CT449	Generators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	30/11/2021	30/11/2021
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	30/11/2021	30/11/2021
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	30/11/2021	30/11/2021
					CT981	Single use	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	30/11/2021	30/11/2021
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	30/11/2021	30/11/2021
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	30/11/2021	30/11/2021
					CT983	Surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	30/11/2021	30/11/2021
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	30/11/2021	30/11/2021
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	30/11/2021	30/11/2021
7343	2022	All other reportable incidents	35852	Single-chamber implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem	13/10/2022	13/10/2022
									A070504	Premature Discharge of Battery	13/10/2022	13/10/2022
							F08	Hospitalization or Prolonged Hospitalization	A0705	Battery Problem	13/10/2022	13/10/2022
									A070504	Premature Discharge of Battery	13/10/2022	13/10/2022
							F1903	Device Explantation	A0705	Battery Problem	13/10/2022	13/10/2022
									A070504	Premature Discharge of Battery	13/10/2022	13/10/2022
7344	2022	All other reportable incidents	17882	Physiologic-monitoring defibrillation system	CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070803	Failure to Power Up		
							F27	Problem Identified During Non-Clinical Procedure	A070803	Failure to Power Up		
7345	2022	All other reportable incidents	47265	Dual-chamber implantable pacemaker, rate-responsive	CT141	Electrodes/Leads and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	04/05/2022	04/05/2022
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	04/05/2022	04/05/2022
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	04/05/2022	04/05/2022
					CT145	Cardiovascular devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	04/05/2022	04/05/2022
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	04/05/2022	04/05/2022
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	04/05/2022	04/05/2022
					CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	04/05/2022	04/05/2022
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	04/05/2022	04/05/2022
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	04/05/2022	04/05/2022
					CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	04/05/2022	04/05/2022
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	04/05/2022	04/05/2022
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	04/05/2022	04/05/2022
					CT233	Surgical	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	04/05/2022	04/05/2022
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	04/05/2022	04/05/2022
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	04/05/2022	04/05/2022
					CT2406	Implantable	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	04/05/2022	04/05/2022
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	04/05/2022	04/05/2022
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	04/05/2022	04/05/2022
					CT2407	Active implantable	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	04/05/2022	04/05/2022
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	04/05/2022	04/05/2022
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	04/05/2022	04/05/2022
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	04/05/2022	04/05/2022
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	04/05/2022	04/05/2022
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	04/05/2022	04/05/2022
					CT2727	Active implant devices/ancillaries	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	04/05/2022	04/05/2022
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	04/05/2022	04/05/2022
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	04/05/2022	04/05/2022
					CT2750	Implantable cardiac electrical stimulators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	04/05/2022	04/05/2022
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	04/05/2022	04/05/2022
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	04/05/2022	04/05/2022

					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	04/05/2022	04/05/2022
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	04/05/2022	04/05/2022
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	04/05/2022	04/05/2022
					CT2985	Active	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	04/05/2022	04/05/2022
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	04/05/2022	04/05/2022
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	04/05/2022	04/05/2022
					CT3001	Implantable metal	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	04/05/2022	04/05/2022
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	04/05/2022	04/05/2022
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	04/05/2022	04/05/2022
					CT3002	Implantable synthetic polymer	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	04/05/2022	04/05/2022
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	04/05/2022	04/05/2022
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	04/05/2022	04/05/2022
					CT3006	Implantable inorganic material	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	04/05/2022	04/05/2022
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	04/05/2022	04/05/2022
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	04/05/2022	04/05/2022
					CT314	Electrically powered	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	04/05/2022	04/05/2022
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	04/05/2022	04/05/2022
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	04/05/2022	04/05/2022
					CT321	Long-term surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	04/05/2022	04/05/2022
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	04/05/2022	04/05/2022
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	04/05/2022	04/05/2022
					CT449	Generators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	04/05/2022	04/05/2022
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	04/05/2022	04/05/2022
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	04/05/2022	04/05/2022
					CT981	Single use	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	04/05/2022	04/05/2022
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	04/05/2022	04/05/2022
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	04/05/2022	04/05/2022
					CT983	Surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	04/05/2022	04/05/2022
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	04/05/2022	04/05/2022
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	04/05/2022	04/05/2022
7346	2022	All other reportable incidents	35852	Single-chamber implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem	15/11/2022	15/11/2022
									A070504	Premature Discharge of Battery	15/11/2022	15/11/2022
							F08	Hospitalization or Prolonged Hospitalization	A0705	Battery Problem	15/11/2022	15/11/2022
									A070504	Premature Discharge of Battery	15/11/2022	15/11/2022
							F1903	Device Explantation	A0705	Battery Problem	15/11/2022	15/11/2022
									A070504	Premature Discharge of Battery	15/11/2022	15/11/2022
7347	2022	All other reportable incidents	35852	Single-chamber implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem	20/10/2022	20/10/2022
									A070504	Premature Discharge of Battery	20/10/2022	20/10/2022
							F08	Hospitalization or Prolonged Hospitalization	A0705	Battery Problem	20/10/2022	20/10/2022
									A070504	Premature Discharge of Battery	20/10/2022	20/10/2022
							F1905	Device Revision or Replacement	A0705	Battery Problem	20/10/2022	20/10/2022
									A070504	Premature Discharge of Battery	20/10/2022	20/10/2022
7348	2022	All other reportable incidents	35852	Single-chamber implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	25/08/2022	25/08/2022
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	25/08/2022	25/08/2022
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	25/08/2022	25/08/2022
7349	2022	All other reportable incidents	35852	Single-chamber implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	30/08/2022	30/08/2022
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	30/08/2022	30/08/2022
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	30/08/2022	30/08/2022
7350	2022	All other reportable incidents	47263	Cardiac resynchronization therapy implantable pacemaker	CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	26/07/2022	26/07/2022
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery	26/07/2022	26/07/2022
					CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	26/07/2022	26/07/2022
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery	26/07/2022	26/07/2022
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	26/07/2022	26/07/2022
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery	26/07/2022	26/07/2022
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	26/07/2022	26/07/2022
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery	26/07/2022	26/07/2022
7351	2021	All other reportable incidents	35223	Endocardial/interventricular septal pacing lead	CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A0401	Break		
									A040101	Fracture		
									A070901	Decreased Sensitivity		
									A070909	Oversensing		
									A072001	Electromagnetic Interference		
									A090801	Signal Artifact/Noise		
							F26	No Health Consequences or Impact	A0401	Break		
									A040101	Fracture		
									A070901	Decreased Sensitivity		
									A070909	Oversensing		
									A072001	Electromagnetic Interference		
									A090801	Signal Artifact/Noise		
7352	2021	All other reportable incidents	47265	Dual-chamber implantable pacemaker, rate-responsive	CT141	Electrodes/Leads and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070909	Oversensing		
									A090801	Signal Artifact/Noise		
							F2203	Imaging Required	A070909	Oversensing		
									A090801	Signal Artifact/Noise		
							F26	No Health Consequences or Impact	A070909	Oversensing		
									A090801	Signal Artifact/Noise		
					CT145	Cardiovascular devices	E2403	No Clinical Signs, Symptoms or Conditions	A070909	Oversensing		
									A090801	Signal Artifact/Noise		
							F2203	Imaging Required	A070909	Oversensing		
									A090801	Signal Artifact/Noise		
							F26	No Health Consequences or Impact	A070909	Oversensing		
									A090801	Signal Artifact/Noise		
					CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070909	Oversensing		
									A090801	Signal Artifact/Noise		
							F2203	Imaging Required	A070909	Oversensing		
									A090801	Signal Artifact/Noise		
							F26	No Health Consequences or Impact	A070909	Oversensing		
									A090801	Signal Artifact/Noise		
					CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A070909	Oversensing		
									A090801	Signal Artifact/Noise		

		F2203	Imaging Required	A070909	Oversensing		
				A090801	Signal Artifact/Noise		
		F26	No Health Consequences or Impact	A070909	Oversensing		
				A090801	Signal Artifact/Noise		
CT233	Surgical	E2403	No Clinical Signs, Symptoms or Conditions	A070909	Oversensing		
				A090801	Signal Artifact/Noise		
		F2203	Imaging Required	A070909	Oversensing		
				A090801	Signal Artifact/Noise		
		F26	No Health Consequences or Impact	A070909	Oversensing		
				A090801	Signal Artifact/Noise		
CT2406	Implantable	E2403	No Clinical Signs, Symptoms or Conditions	A070909	Oversensing		
				A090801	Signal Artifact/Noise		
		F2203	Imaging Required	A070909	Oversensing		
				A090801	Signal Artifact/Noise		
		F26	No Health Consequences or Impact	A070909	Oversensing		
				A090801	Signal Artifact/Noise		
CT2407	Active implantable	E2403	No Clinical Signs, Symptoms or Conditions	A070909	Oversensing		
				A090801	Signal Artifact/Noise		
		F2203	Imaging Required	A070909	Oversensing		
				A090801	Signal Artifact/Noise		
		F26	No Health Consequences or Impact	A070909	Oversensing		
				A090801	Signal Artifact/Noise		
CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A070909	Oversensing		
				A090801	Signal Artifact/Noise		
		F2203	Imaging Required	A070909	Oversensing		
				A090801	Signal Artifact/Noise		
		F26	No Health Consequences or Impact	A070909	Oversensing		
				A090801	Signal Artifact/Noise		
CT2727	Active implant devices/ancillaries	E2403	No Clinical Signs, Symptoms or Conditions	A070909	Oversensing		
				A090801	Signal Artifact/Noise		
		F2203	Imaging Required	A070909	Oversensing		
				A090801	Signal Artifact/Noise		
		F26	No Health Consequences or Impact	A070909	Oversensing		
				A090801	Signal Artifact/Noise		
CT2750	Implantable cardiac electrical stimulators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070909	Oversensing		
				A090801	Signal Artifact/Noise		
		F2203	Imaging Required	A070909	Oversensing		
				A090801	Signal Artifact/Noise		
		F26	No Health Consequences or Impact	A070909	Oversensing		
				A090801	Signal Artifact/Noise		
CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A070909	Oversensing		
				A090801	Signal Artifact/Noise		
		F2203	Imaging Required	A070909	Oversensing		
				A090801	Signal Artifact/Noise		
		F26	No Health Consequences or Impact	A070909	Oversensing		
				A090801	Signal Artifact/Noise		
CT2985	Active	E2403	No Clinical Signs, Symptoms or Conditions	A070909	Oversensing		
				A090801	Signal Artifact/Noise		
		F2203	Imaging Required	A070909	Oversensing		
				A090801	Signal Artifact/Noise		
		F26	No Health Consequences or Impact	A070909	Oversensing		
				A090801	Signal Artifact/Noise		
CT3001	Implantable metal	E2403	No Clinical Signs, Symptoms or Conditions	A070909	Oversensing		
				A090801	Signal Artifact/Noise		
		F2203	Imaging Required	A070909	Oversensing		
				A090801	Signal Artifact/Noise		
		F26	No Health Consequences or Impact	A070909	Oversensing		
				A090801	Signal Artifact/Noise		
CT3002	Implantable synthetic polymer	E2403	No Clinical Signs, Symptoms or Conditions	A070909	Oversensing		
				A090801	Signal Artifact/Noise		
		F2203	Imaging Required	A070909	Oversensing		
				A090801	Signal Artifact/Noise		
		F26	No Health Consequences or Impact	A070909	Oversensing		
				A090801	Signal Artifact/Noise		
CT3006	Implantable inorganic material	E2403	No Clinical Signs, Symptoms or Conditions	A070909	Oversensing		
				A090801	Signal Artifact/Noise		
		F2203	Imaging Required	A070909	Oversensing		
				A090801	Signal Artifact/Noise		
		F26	No Health Consequences or Impact	A070909	Oversensing		
				A090801	Signal Artifact/Noise		
CT314	Electrically powered	E2403	No Clinical Signs, Symptoms or Conditions	A070909	Oversensing		
				A090801	Signal Artifact/Noise		
		F2203	Imaging Required	A070909	Oversensing		
				A090801	Signal Artifact/Noise		
		F26	No Health Consequences or Impact	A070909	Oversensing		
				A090801	Signal Artifact/Noise		
CT321	Long-term surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A070909	Oversensing		
				A090801	Signal Artifact/Noise		
		F2203	Imaging Required	A070909	Oversensing		
				A090801	Signal Artifact/Noise		
		F26	No Health Consequences or Impact	A070909	Oversensing		
				A090801	Signal Artifact/Noise		
CT449	Generators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070909	Oversensing		
				A090801	Signal Artifact/Noise		
		F2203	Imaging Required	A070909	Oversensing		
				A090801	Signal Artifact/Noise		
		F26	No Health Consequences or Impact	A070909	Oversensing		
				A090801	Signal Artifact/Noise		

					CT981	Single use	E2403	No Clinical Signs, Symptoms or Conditions	A070909	Oversensing		
									A090801	Signal Artifact/Noise		
							F2203	Imaging Required	A070909	Oversensing		
									A090801	Signal Artifact/Noise		
							F26	No Health Consequences or Impact	A070909	Oversensing		
									A090801	Signal Artifact/Noise		
					CT983	Surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A070909	Oversensing		
									A090801	Signal Artifact/Noise		
							F2203	Imaging Required	A070909	Oversensing		
									A090801	Signal Artifact/Noise		
							F26	No Health Consequences or Impact	A070909	Oversensing		
									A090801	Signal Artifact/Noise		
7353	2022	All other reportable incidents	47270	Cardiac resynchronization therapy implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A070909	Oversensing		
									A071302	Inappropriate/Inadequate Shock/Stimulation		
									A0908	Incorrect, Inadequate or Imprecise Result or Readings		
							F12	Serious Injury/ Illness/ Impairment	A070909	Oversensing		
									A071302	Inappropriate/Inadequate Shock/Stimulation		
									A0908	Incorrect, Inadequate or Imprecise Result or Readings		
7354	2022	All other reportable incidents	47205	Cardiac pulse generator programmer	CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A090805	Incorrect Measurement		
							F23	Unexpected Medical Intervention	A090805	Incorrect Measurement		
7355	2022	All other reportable incidents	47265	Dual-chamber implantable pacemaker, rate-responsive	CT141	Electrodes/Leads and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A0706	Charging Problem	28/07/2022	28/07/2022
									A071205	Pacemaker Found in Back-Up Mode	28/07/2022	28/07/2022
							F1901	Additional Surgery	A0706	Charging Problem	28/07/2022	28/07/2022
									A071205	Pacemaker Found in Back-Up Mode	28/07/2022	28/07/2022
					CT145	Cardiovascular devices	E2403	No Clinical Signs, Symptoms or Conditions	A0706	Charging Problem	28/07/2022	28/07/2022
									A071205	Pacemaker Found in Back-Up Mode	28/07/2022	28/07/2022
							F1901	Additional Surgery	A0706	Charging Problem	28/07/2022	28/07/2022
									A071205	Pacemaker Found in Back-Up Mode	28/07/2022	28/07/2022
					CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A0706	Charging Problem	28/07/2022	28/07/2022
									A071205	Pacemaker Found in Back-Up Mode	28/07/2022	28/07/2022
							F1901	Additional Surgery	A0706	Charging Problem	28/07/2022	28/07/2022
									A071205	Pacemaker Found in Back-Up Mode	28/07/2022	28/07/2022
					CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A0706	Charging Problem	28/07/2022	28/07/2022
									A071205	Pacemaker Found in Back-Up Mode	28/07/2022	28/07/2022
							F1901	Additional Surgery	A0706	Charging Problem	28/07/2022	28/07/2022
									A071205	Pacemaker Found in Back-Up Mode	28/07/2022	28/07/2022
					CT233	Surgical	E2403	No Clinical Signs, Symptoms or Conditions	A0706	Charging Problem	28/07/2022	28/07/2022
									A071205	Pacemaker Found in Back-Up Mode	28/07/2022	28/07/2022
							F1901	Additional Surgery	A0706	Charging Problem	28/07/2022	28/07/2022
									A071205	Pacemaker Found in Back-Up Mode	28/07/2022	28/07/2022
					CT2406	Implantable	E2403	No Clinical Signs, Symptoms or Conditions	A0706	Charging Problem	28/07/2022	28/07/2022
									A071205	Pacemaker Found in Back-Up Mode	28/07/2022	28/07/2022
							F1901	Additional Surgery	A0706	Charging Problem	28/07/2022	28/07/2022
									A071205	Pacemaker Found in Back-Up Mode	28/07/2022	28/07/2022
					CT2407	Active implantable	E2403	No Clinical Signs, Symptoms or Conditions	A0706	Charging Problem	28/07/2022	28/07/2022
									A071205	Pacemaker Found in Back-Up Mode	28/07/2022	28/07/2022
							F1901	Additional Surgery	A0706	Charging Problem	28/07/2022	28/07/2022
									A071205	Pacemaker Found in Back-Up Mode	28/07/2022	28/07/2022
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A0706	Charging Problem	28/07/2022	28/07/2022
									A071205	Pacemaker Found in Back-Up Mode	28/07/2022	28/07/2022
							F1901	Additional Surgery	A0706	Charging Problem	28/07/2022	28/07/2022
									A071205	Pacemaker Found in Back-Up Mode	28/07/2022	28/07/2022
					CT2727	Active implant devices/ancillaries	E2403	No Clinical Signs, Symptoms or Conditions	A0706	Charging Problem	28/07/2022	28/07/2022
									A071205	Pacemaker Found in Back-Up Mode	28/07/2022	28/07/2022
							F1901	Additional Surgery	A0706	Charging Problem	28/07/2022	28/07/2022
									A071205	Pacemaker Found in Back-Up Mode	28/07/2022	28/07/2022
					CT2750	Implantable cardiac electrical stimulators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A0706	Charging Problem	28/07/2022	28/07/2022
									A071205	Pacemaker Found in Back-Up Mode	28/07/2022	28/07/2022
							F1901	Additional Surgery	A0706	Charging Problem	28/07/2022	28/07/2022
									A071205	Pacemaker Found in Back-Up Mode	28/07/2022	28/07/2022
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A0706	Charging Problem	28/07/2022	28/07/2022
									A071205	Pacemaker Found in Back-Up Mode	28/07/2022	28/07/2022
							F1901	Additional Surgery	A0706	Charging Problem	28/07/2022	28/07/2022
									A071205	Pacemaker Found in Back-Up Mode	28/07/2022	28/07/2022
					CT2985	Active	E2403	No Clinical Signs, Symptoms or Conditions	A0706	Charging Problem	28/07/2022	28/07/2022
									A071205	Pacemaker Found in Back-Up Mode	28/07/2022	28/07/2022
							F1901	Additional Surgery	A0706	Charging Problem	28/07/2022	28/07/2022
									A071205	Pacemaker Found in Back-Up Mode	28/07/2022	28/07/2022
					CT3001	Implantable metal	E2403	No Clinical Signs, Symptoms or Conditions	A0706	Charging Problem	28/07/2022	28/07/2022
									A071205	Pacemaker Found in Back-Up Mode	28/07/2022	28/07/2022
							F1901	Additional Surgery	A0706	Charging Problem	28/07/2022	28/07/2022
									A071205	Pacemaker Found in Back-Up Mode	28/07/2022	28/07/2022
					CT3002	Implantable synthetic polymer	E2403	No Clinical Signs, Symptoms or Conditions	A0706	Charging Problem	28/07/2022	28/07/2022
									A071205	Pacemaker Found in Back-Up Mode	28/07/2022	28/07/2022
							F1901	Additional Surgery	A0706	Charging Problem	28/07/2022	28/07/2022
									A071205	Pacemaker Found in Back-Up Mode	28/07/2022	28/07/2022
					CT3006	Implantable inorganic material	E2403	No Clinical Signs, Symptoms or Conditions	A0706	Charging Problem	28/07/2022	28/07/2022
									A071205	Pacemaker Found in Back-Up Mode	28/07/2022	28/07/2022
							F1901	Additional Surgery	A0706	Charging Problem	28/07/2022	28/07/2022
									A071205	Pacemaker Found in Back-Up Mode	28/07/2022	28/07/2022
					CT314	Electrically powered	E2403	No Clinical Signs, Symptoms or Conditions	A0706	Charging Problem	28/07/2022	28/07/2022
									A071205	Pacemaker Found in Back-Up Mode	28/07/2022	28/07/2022
							F1901	Additional Surgery	A0706	Charging Problem	28/07/2022	28/07/2022
									A071205	Pacemaker Found in Back-Up Mode	28/07/2022	28/07/2022
					CT321	Long-term surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A0706	Charging Problem	28/07/2022	28/07/2022
									A071205	Pacemaker Found in Back-Up Mode	28/07/2022	28/07/2022

							F1901	Additional Surgery	A0706	Charging Problem	28/07/2022	28/07/2022
									A071205	Pacemaker Found in Back-Up Mode	28/07/2022	28/07/2022
					CT449	Generators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A0706	Charging Problem	28/07/2022	28/07/2022
									A071205	Pacemaker Found in Back-Up Mode	28/07/2022	28/07/2022
							F1901	Additional Surgery	A0706	Charging Problem	28/07/2022	28/07/2022
									A071205	Pacemaker Found in Back-Up Mode	28/07/2022	28/07/2022
					CT981	Single use	E2403	No Clinical Signs, Symptoms or Conditions	A0706	Charging Problem	28/07/2022	28/07/2022
									A071205	Pacemaker Found in Back-Up Mode	28/07/2022	28/07/2022
							F1901	Additional Surgery	A0706	Charging Problem	28/07/2022	28/07/2022
									A071205	Pacemaker Found in Back-Up Mode	28/07/2022	28/07/2022
					CT983	Surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A0706	Charging Problem	28/07/2022	28/07/2022
									A071205	Pacemaker Found in Back-Up Mode	28/07/2022	28/07/2022
							F1901	Additional Surgery	A0706	Charging Problem	28/07/2022	28/07/2022
									A071205	Pacemaker Found in Back-Up Mode	28/07/2022	28/07/2022
7356	2021	All other reportable incidents	47270	Cardiac resynchronization therapy implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E2311	Discomfort	A070909	Oversensing		
									A071204	No Pacing		
							E2403	No Clinical Signs, Symptoms or Conditions	A070909	Oversensing		
									A071204	No Pacing		
							F23	Unexpected Medical Intervention	A070909	Oversensing		
									A071204	No Pacing		
7357	2022	All other reportable incidents	17882	Physiologic-monitoring defibrillation system	CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A07	Electrical /Electronic Property Problem		
							F27	Problem Identified During Non-Clinical Procedure	A07	Electrical /Electronic Property Problem		
7358	2021	All other reportable incidents	35852	Single-chamber implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A0908	Incorrect, Inadequate or Imprecise Result or Readings		
									A160503	Inappropriate or Unexpected Reset		
							F26	No Health Consequences or Impact	A0908	Incorrect, Inadequate or Imprecise Result or Readings		
									A160503	Inappropriate or Unexpected Reset		
7359	2021	All other reportable incidents	47267	Single-chamber implantable pacemaker, rate-responsive	CT141	Electrodes/Leads and associated devices	E0112	Dizziness	A070101	Failure to Capture		
									A072201	High Impedance		
									A1301	Failure to Read Input Signal		
							E2403	No Clinical Signs, Symptoms or Conditions	A070101	Failure to Capture		
									A072201	High Impedance		
									A1301	Failure to Read Input Signal		
							F15	Recognized Device or Procedural Complication	A070101	Failure to Capture		
									A072201	High Impedance		
									A1301	Failure to Read Input Signal		
					CT145	Cardiovascular devices	E0112	Dizziness	A070101	Failure to Capture		
									A072201	High Impedance		
									A1301	Failure to Read Input Signal		
							E2403	No Clinical Signs, Symptoms or Conditions	A070101	Failure to Capture		
									A072201	High Impedance		
									A1301	Failure to Read Input Signal		
							F15	Recognized Device or Procedural Complication	A070101	Failure to Capture		
									A072201	High Impedance		
									A1301	Failure to Read Input Signal		
					CT1684	Cardiac pacemakers and associated devices	E0112	Dizziness	A070101	Failure to Capture		
									A072201	High Impedance		
									A1301	Failure to Read Input Signal		
							E2403	No Clinical Signs, Symptoms or Conditions	A070101	Failure to Capture		
									A072201	High Impedance		
									A1301	Failure to Read Input Signal		
							F15	Recognized Device or Procedural Complication	A070101	Failure to Capture		
									A072201	High Impedance		
									A1301	Failure to Read Input Signal		
					CT1685	Cardiac pacemakers	E0112	Dizziness	A070101	Failure to Capture		
									A072201	High Impedance		
									A1301	Failure to Read Input Signal		
							E2403	No Clinical Signs, Symptoms or Conditions	A070101	Failure to Capture		
									A072201	High Impedance		
									A1301	Failure to Read Input Signal		
							F15	Recognized Device or Procedural Complication	A070101	Failure to Capture		
									A072201	High Impedance		
									A1301	Failure to Read Input Signal		
					CT233	Surgical	E0112	Dizziness	A070101	Failure to Capture		
									A072201	High Impedance		
									A1301	Failure to Read Input Signal		
							E2403	No Clinical Signs, Symptoms or Conditions	A070101	Failure to Capture		
									A072201	High Impedance		
									A1301	Failure to Read Input Signal		
							F15	Recognized Device or Procedural Complication	A070101	Failure to Capture		
									A072201	High Impedance		
									A1301	Failure to Read Input Signal		
					CT2406	Implantable	E0112	Dizziness	A070101	Failure to Capture		
									A072201	High Impedance		
									A1301	Failure to Read Input Signal		
							E2403	No Clinical Signs, Symptoms or Conditions	A070101	Failure to Capture		
									A072201	High Impedance		
									A1301	Failure to Read Input Signal		
							F15	Recognized Device or Procedural Complication	A070101	Failure to Capture		
									A072201	High Impedance		
									A1301	Failure to Read Input Signal		
					CT2407	Active implantable	E0112	Dizziness	A070101	Failure to Capture		
									A072201	High Impedance		
									A1301	Failure to Read Input Signal		
							E2403	No Clinical Signs, Symptoms or Conditions	A070101	Failure to Capture		
									A072201	High Impedance		
									A1301	Failure to Read Input Signal		

							F15	Recognized Device or Procedural Complication	A070101	Failure to Capture		
									A072201	High Impedance		
									A1301	Failure to Read Input Signal		
					CT257	Implantable cardiac pacemakers	E0112	Dizziness	A070101	Failure to Capture		
									A072201	High Impedance		
									A1301	Failure to Read Input Signal		
							E2403	No Clinical Signs, Symptoms or Conditions	A070101	Failure to Capture		
									A072201	High Impedance		
									A1301	Failure to Read Input Signal		
							F15	Recognized Device or Procedural Complication	A070101	Failure to Capture		
									A072201	High Impedance		
									A1301	Failure to Read Input Signal		
					CT2727	Active implant devices/ancillaries	E0112	Dizziness	A070101	Failure to Capture		
									A072201	High Impedance		
									A1301	Failure to Read Input Signal		
							E2403	No Clinical Signs, Symptoms or Conditions	A070101	Failure to Capture		
									A072201	High Impedance		
									A1301	Failure to Read Input Signal		
							F15	Recognized Device or Procedural Complication	A070101	Failure to Capture		
									A072201	High Impedance		
									A1301	Failure to Read Input Signal		
					CT2750	Implantable cardiac electrical stimulators and associated devices	E0112	Dizziness	A070101	Failure to Capture		
									A072201	High Impedance		
									A1301	Failure to Read Input Signal		
							E2403	No Clinical Signs, Symptoms or Conditions	A070101	Failure to Capture		
									A072201	High Impedance		
									A1301	Failure to Read Input Signal		
							F15	Recognized Device or Procedural Complication	A070101	Failure to Capture		
									A072201	High Impedance		
									A1301	Failure to Read Input Signal		
					CT2866	Implantable cardiac electrical stimulators	E0112	Dizziness	A070101	Failure to Capture		
									A072201	High Impedance		
									A1301	Failure to Read Input Signal		
							E2403	No Clinical Signs, Symptoms or Conditions	A070101	Failure to Capture		
									A072201	High Impedance		
									A1301	Failure to Read Input Signal		
							F15	Recognized Device or Procedural Complication	A070101	Failure to Capture		
									A072201	High Impedance		
									A1301	Failure to Read Input Signal		
					CT2985	Active	E0112	Dizziness	A070101	Failure to Capture		
									A072201	High Impedance		
									A1301	Failure to Read Input Signal		
							E2403	No Clinical Signs, Symptoms or Conditions	A070101	Failure to Capture		
									A072201	High Impedance		
									A1301	Failure to Read Input Signal		
							F15	Recognized Device or Procedural Complication	A070101	Failure to Capture		
									A072201	High Impedance		
									A1301	Failure to Read Input Signal		
					CT3001	Implantable metal	E0112	Dizziness	A070101	Failure to Capture		
									A072201	High Impedance		
									A1301	Failure to Read Input Signal		
							E2403	No Clinical Signs, Symptoms or Conditions	A070101	Failure to Capture		
									A072201	High Impedance		
									A1301	Failure to Read Input Signal		
							F15	Recognized Device or Procedural Complication	A070101	Failure to Capture		
									A072201	High Impedance		
									A1301	Failure to Read Input Signal		
					CT3002	Implantable synthetic polymer	E0112	Dizziness	A070101	Failure to Capture		
									A072201	High Impedance		
									A1301	Failure to Read Input Signal		
							E2403	No Clinical Signs, Symptoms or Conditions	A070101	Failure to Capture		
									A072201	High Impedance		
									A1301	Failure to Read Input Signal		
							F15	Recognized Device or Procedural Complication	A070101	Failure to Capture		
									A072201	High Impedance		
									A1301	Failure to Read Input Signal		
					CT3006	Implantable inorganic material	E0112	Dizziness	A070101	Failure to Capture		
									A072201	High Impedance		
									A1301	Failure to Read Input Signal		
							E2403	No Clinical Signs, Symptoms or Conditions	A070101	Failure to Capture		
									A072201	High Impedance		
									A1301	Failure to Read Input Signal		
							F15	Recognized Device or Procedural Complication	A070101	Failure to Capture		
									A072201	High Impedance		
									A1301	Failure to Read Input Signal		
					CT314	Electrically powered	E0112	Dizziness	A070101	Failure to Capture		
									A072201	High Impedance		
									A1301	Failure to Read Input Signal		
							E2403	No Clinical Signs, Symptoms or Conditions	A070101	Failure to Capture		
									A072201	High Impedance		
									A1301	Failure to Read Input Signal		
							F15	Recognized Device or Procedural Complication	A070101	Failure to Capture		
									A072201	High Impedance		
									A1301	Failure to Read Input Signal		
					CT321	Long-term surgical invasive	E0112	Dizziness	A070101	Failure to Capture		
									A072201	High Impedance		
									A1301	Failure to Read Input Signal		
							E2403	No Clinical Signs, Symptoms or Conditions	A070101	Failure to Capture		

								A072201	High Impedance			
								A1301	Failure to Read Input Signal			
						F15	Recognized Device or Procedural Complication	A070101	Failure to Capture			
								A072201	High Impedance			
								A1301	Failure to Read Input Signal			
					CT449	Generators and associated devices	E0112	Dizziness	A070101	Failure to Capture		
									A072201	High Impedance		
									A1301	Failure to Read Input Signal		
							E2403	No Clinical Signs, Symptoms or Conditions	A070101	Failure to Capture		
									A072201	High Impedance		
									A1301	Failure to Read Input Signal		
							F15	Recognized Device or Procedural Complication	A070101	Failure to Capture		
									A072201	High Impedance		
									A1301	Failure to Read Input Signal		
					CT981	Single use	E0112	Dizziness	A070101	Failure to Capture		
									A072201	High Impedance		
									A1301	Failure to Read Input Signal		
							E2403	No Clinical Signs, Symptoms or Conditions	A070101	Failure to Capture		
									A072201	High Impedance		
									A1301	Failure to Read Input Signal		
							F15	Recognized Device or Procedural Complication	A070101	Failure to Capture		
									A072201	High Impedance		
									A1301	Failure to Read Input Signal		
					CT983	Surgical invasive	E0112	Dizziness	A070101	Failure to Capture		
									A072201	High Impedance		
									A1301	Failure to Read Input Signal		
							E2403	No Clinical Signs, Symptoms or Conditions	A070101	Failure to Capture		
									A072201	High Impedance		
									A1301	Failure to Read Input Signal		
							F15	Recognized Device or Procedural Complication	A070101	Failure to Capture		
									A072201	High Impedance		
									A1301	Failure to Read Input Signal		
7360	2022	All other reportable incidents	17882	Physiologic-monitoring defibrillation system	CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070803	Failure to Power Up		
							F27	Problem Identified During Non-Clinical Procedure	A070803	Failure to Power Up		
7361	2021	All other reportable incidents	47265	Dual-chamber implantable pacemaker, rate-responsive	CT141	Electrodes/Leads and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	07/12/2021	07/12/2021
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	07/12/2021	07/12/2021
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	07/12/2021	07/12/2021
					CT145	Cardiovascular devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	07/12/2021	07/12/2021
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	07/12/2021	07/12/2021
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	07/12/2021	07/12/2021
					CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	07/12/2021	07/12/2021
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	07/12/2021	07/12/2021
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	07/12/2021	07/12/2021
					CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	07/12/2021	07/12/2021
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	07/12/2021	07/12/2021
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	07/12/2021	07/12/2021
					CT233	Surgical	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	07/12/2021	07/12/2021
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	07/12/2021	07/12/2021
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	07/12/2021	07/12/2021
					CT2406	Implantable	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	07/12/2021	07/12/2021
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	07/12/2021	07/12/2021
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	07/12/2021	07/12/2021
					CT2407	Active implantable	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	07/12/2021	07/12/2021
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	07/12/2021	07/12/2021
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	07/12/2021	07/12/2021
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	07/12/2021	07/12/2021
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	07/12/2021	07/12/2021
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	07/12/2021	07/12/2021
					CT2727	Active implant devices/ancillaries	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	07/12/2021	07/12/2021
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	07/12/2021	07/12/2021
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	07/12/2021	07/12/2021
					CT2750	Implantable cardiac electrical stimulators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	07/12/2021	07/12/2021
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	07/12/2021	07/12/2021
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	07/12/2021	07/12/2021
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	07/12/2021	07/12/2021
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	07/12/2021	07/12/2021
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	07/12/2021	07/12/2021
					CT2985	Active	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	07/12/2021	07/12/2021
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	07/12/2021	07/12/2021
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	07/12/2021	07/12/2021
					CT3001	Implantable metal	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	07/12/2021	07/12/2021
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	07/12/2021	07/12/2021
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	07/12/2021	07/12/2021
					CT3002	Implantable synthetic polymer	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	07/12/2021	07/12/2021
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	07/12/2021	07/12/2021
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	07/12/2021	07/12/2021
					CT3006	Implantable inorganic material	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	07/12/2021	07/12/2021
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	07/12/2021	07/12/2021
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	07/12/2021	07/12/2021
					CT314	Electrically powered	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	07/12/2021	07/12/2021
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	07/12/2021	07/12/2021
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	07/12/2021	07/12/2021
					CT321	Long-term surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	07/12/2021	07/12/2021
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	07/12/2021	07/12/2021
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	07/12/2021	07/12/2021
					CT449	Generators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	07/12/2021	07/12/2021
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	07/12/2021	07/12/2021
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	07/12/2021	07/12/2021

					CT981	Single use	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	07/12/2021	07/12/2021
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	07/12/2021	07/12/2021
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	07/12/2021	07/12/2021
					CT983	Surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	07/12/2021	07/12/2021
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	07/12/2021	07/12/2021
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	07/12/2021	07/12/2021
7362	2022	All other reportable incidents	35223	Endocardial/interventricular septal pacing lead	CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A072202	Low Impedance		
							F26	No Health Consequences or Impact	A072202	Low Impedance		
7363	2022	All other reportable incidents	47265	Dual-chamber implantable pacemaker, rate-responsive	CT141	Electrodes/Leads and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT145	Cardiovascular devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT233	Surgical	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT2406	Implantable	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT2407	Active implantable	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT2727	Active implant devices/ancillaries	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT2750	Implantable cardiac electrical stimulators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT2985	Active	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT3001	Implantable metal	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT3002	Implantable synthetic polymer	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT3006	Implantable inorganic material	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT314	Electrically powered	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT321	Long-term surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT449	Generators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT981	Single use	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT983	Surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
7364	2022	All other reportable incidents	47265	Dual-chamber implantable pacemaker, rate-responsive	CT141	Electrodes/Leads and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A071102	Failure to Interrogate	03/08/2022	03/08/2022
									A071204	No Pacing	03/08/2022	03/08/2022
							F1901	Additional Surgery	A071102	Failure to Interrogate	03/08/2022	03/08/2022
									A071204	No Pacing	03/08/2022	03/08/2022
					CT145	Cardiovascular devices	E2403	No Clinical Signs, Symptoms or Conditions	A071102	Failure to Interrogate	03/08/2022	03/08/2022
									A071204	No Pacing	03/08/2022	03/08/2022
							F1901	Additional Surgery	A071102	Failure to Interrogate	03/08/2022	03/08/2022
									A071204	No Pacing	03/08/2022	03/08/2022
					CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A071102	Failure to Interrogate	03/08/2022	03/08/2022
									A071204	No Pacing	03/08/2022	03/08/2022
							F1901	Additional Surgery	A071102	Failure to Interrogate	03/08/2022	03/08/2022
									A071204	No Pacing	03/08/2022	03/08/2022
					CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A071102	Failure to Interrogate	03/08/2022	03/08/2022
									A071204	No Pacing	03/08/2022	03/08/2022
							F1901	Additional Surgery	A071102	Failure to Interrogate	03/08/2022	03/08/2022
									A071204	No Pacing	03/08/2022	03/08/2022
					CT233	Surgical	E2403	No Clinical Signs, Symptoms or Conditions	A071102	Failure to Interrogate	03/08/2022	03/08/2022
									A071204	No Pacing	03/08/2022	03/08/2022
							F1901	Additional Surgery	A071102	Failure to Interrogate	03/08/2022	03/08/2022
									A071204	No Pacing	03/08/2022	03/08/2022
					CT2406	Implantable	E2403	No Clinical Signs, Symptoms or Conditions	A071102	Failure to Interrogate	03/08/2022	03/08/2022
									A071204	No Pacing	03/08/2022	03/08/2022
							F1901	Additional Surgery	A071102	Failure to Interrogate	03/08/2022	03/08/2022
									A071204	No Pacing	03/08/2022	03/08/2022
					CT2407	Active implantable	E2403	No Clinical Signs, Symptoms or Conditions	A071102	Failure to Interrogate	03/08/2022	03/08/2022
									A071204	No Pacing	03/08/2022	03/08/2022
							F1901	Additional Surgery	A071102	Failure to Interrogate	03/08/2022	03/08/2022
									A071204	No Pacing	03/08/2022	03/08/2022
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A071102	Failure to Interrogate	03/08/2022	03/08/2022
									A071204	No Pacing	03/08/2022	03/08/2022
							F1901	Additional Surgery	A071102	Failure to Interrogate	03/08/2022	03/08/2022
									A071204	No Pacing	03/08/2022	03/08/2022
					CT2727	Active implant devices/ancillaries	E2403	No Clinical Signs, Symptoms or Conditions	A071102	Failure to Interrogate	03/08/2022	03/08/2022
									A071204	No Pacing	03/08/2022	03/08/2022
							F1901	Additional Surgery	A071102	Failure to Interrogate	03/08/2022	03/08/2022
									A071204	No Pacing	03/08/2022	03/08/2022
					CT2750	Implantable cardiac electrical stimulators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A071102	Failure to Interrogate	03/08/2022	03/08/2022
									A071204	No Pacing	03/08/2022	03/08/2022
							F1901	Additional Surgery	A071102	Failure to Interrogate	03/08/2022	03/08/2022
									A071204	No Pacing	03/08/2022	03/08/2022

					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A071102	Failure to Interrogate	03/08/2022	03/08/2022
									A071204	No Pacing	03/08/2022	03/08/2022
							F1901	Additional Surgery	A071102	Failure to Interrogate	03/08/2022	03/08/2022
									A071204	No Pacing	03/08/2022	03/08/2022
					CT2985	Active	E2403	No Clinical Signs, Symptoms or Conditions	A071102	Failure to Interrogate	03/08/2022	03/08/2022
									A071204	No Pacing	03/08/2022	03/08/2022
							F1901	Additional Surgery	A071102	Failure to Interrogate	03/08/2022	03/08/2022
									A071204	No Pacing	03/08/2022	03/08/2022
					CT3001	Implantable metal	E2403	No Clinical Signs, Symptoms or Conditions	A071102	Failure to Interrogate	03/08/2022	03/08/2022
									A071204	No Pacing	03/08/2022	03/08/2022
							F1901	Additional Surgery	A071102	Failure to Interrogate	03/08/2022	03/08/2022
									A071204	No Pacing	03/08/2022	03/08/2022
					CT3002	Implantable synthetic polymer	E2403	No Clinical Signs, Symptoms or Conditions	A071102	Failure to Interrogate	03/08/2022	03/08/2022
									A071204	No Pacing	03/08/2022	03/08/2022
							F1901	Additional Surgery	A071102	Failure to Interrogate	03/08/2022	03/08/2022
									A071204	No Pacing	03/08/2022	03/08/2022
					CT3006	Implantable inorganic material	E2403	No Clinical Signs, Symptoms or Conditions	A071102	Failure to Interrogate	03/08/2022	03/08/2022
									A071204	No Pacing	03/08/2022	03/08/2022
							F1901	Additional Surgery	A071102	Failure to Interrogate	03/08/2022	03/08/2022
									A071204	No Pacing	03/08/2022	03/08/2022
					CT314	Electrically powered	E2403	No Clinical Signs, Symptoms or Conditions	A071102	Failure to Interrogate	03/08/2022	03/08/2022
									A071204	No Pacing	03/08/2022	03/08/2022
							F1901	Additional Surgery	A071102	Failure to Interrogate	03/08/2022	03/08/2022
									A071204	No Pacing	03/08/2022	03/08/2022
					CT321	Long-term surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A071102	Failure to Interrogate	03/08/2022	03/08/2022
									A071204	No Pacing	03/08/2022	03/08/2022
							F1901	Additional Surgery	A071102	Failure to Interrogate	03/08/2022	03/08/2022
									A071204	No Pacing	03/08/2022	03/08/2022
					CT449	Generators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A071102	Failure to Interrogate	03/08/2022	03/08/2022
									A071204	No Pacing	03/08/2022	03/08/2022
							F1901	Additional Surgery	A071102	Failure to Interrogate	03/08/2022	03/08/2022
									A071204	No Pacing	03/08/2022	03/08/2022
					CT981	Single use	E2403	No Clinical Signs, Symptoms or Conditions	A071102	Failure to Interrogate	03/08/2022	03/08/2022
									A071204	No Pacing	03/08/2022	03/08/2022
							F1901	Additional Surgery	A071102	Failure to Interrogate	03/08/2022	03/08/2022
									A071204	No Pacing	03/08/2022	03/08/2022
					CT983	Surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A071102	Failure to Interrogate	03/08/2022	03/08/2022
									A071204	No Pacing	03/08/2022	03/08/2022
							F1901	Additional Surgery	A071102	Failure to Interrogate	03/08/2022	03/08/2022
									A071204	No Pacing	03/08/2022	03/08/2022
7365	2022	All other reportable incidents	37265	Dual-chamber implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E0601	Arrhythmia	A070909	Oversensing		
									A071302	Inappropriate/Inadequate Shock/Stimulation		
							E2403	No Clinical Signs, Symptoms or Conditions	A070909	Oversensing		
									A071302	Inappropriate/Inadequate Shock/Stimulation		
							F12	Serious Injury/ Illness/ Impairment	A070909	Oversensing		
									A071302	Inappropriate/Inadequate Shock/Stimulation		
7366	2021	All other reportable incidents	35852	Single-chamber implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E2104	Electric Shock	A070909	Oversensing		
									A071302	Inappropriate/Inadequate Shock/Stimulation		
							E2403	No Clinical Signs, Symptoms or Conditions	A070909	Oversensing		
									A071302	Inappropriate/Inadequate Shock/Stimulation		
							F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A070909	Oversensing		
									A071302	Inappropriate/Inadequate Shock/Stimulation		
7367	2021	All other reportable incidents	47265	Dual-chamber implantable pacemaker, rate-responsive	CT141	Electrodes/Leads and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem		
							F26	No Health Consequences or Impact	A0705	Battery Problem		
					CT145	Cardiovascular devices	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem		
							F26	No Health Consequences or Impact	A0705	Battery Problem		
					CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem		
							F26	No Health Consequences or Impact	A0705	Battery Problem		
					CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem		
							F26	No Health Consequences or Impact	A0705	Battery Problem		
					CT233	Surgical	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem		
							F26	No Health Consequences or Impact	A0705	Battery Problem		
					CT2406	Implantable	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem		
							F26	No Health Consequences or Impact	A0705	Battery Problem		
					CT2407	Active implantable	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem		
							F26	No Health Consequences or Impact	A0705	Battery Problem		
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem		
							F26	No Health Consequences or Impact	A0705	Battery Problem		
					CT2727	Active implant devices/ancillaries	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem		
							F26	No Health Consequences or Impact	A0705	Battery Problem		
					CT2750	Implantable cardiac electrical stimulators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem		
							F26	No Health Consequences or Impact	A0705	Battery Problem		
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem		
							F26	No Health Consequences or Impact	A0705	Battery Problem		
					CT2985	Active	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem		
							F26	No Health Consequences or Impact	A0705	Battery Problem		
					CT3001	Implantable metal	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem		
							F26	No Health Consequences or Impact	A0705	Battery Problem		
					CT3002	Implantable synthetic polymer	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem		
							F26	No Health Consequences or Impact	A0705	Battery Problem		
					CT3006	Implantable inorganic material	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem		
							F26	No Health Consequences or Impact	A0705	Battery Problem		
					CT314	Electrically powered	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem		
							F26	No Health Consequences or Impact	A0705	Battery Problem		
					CT321	Long-term surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem		
							F26	No Health Consequences or Impact	A0705	Battery Problem		
					CT449	Generators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem		
							F26	No Health Consequences or Impact	A0705	Battery Problem		

					CT981	Single use	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem			
							F26	No Health Consequences or Impact	A0705	Battery Problem			
					CT983	Surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem			
							F26	No Health Consequences or Impact	A0705	Battery Problem			
7368	2021	All other reportable incidents	35852	Single-chamber implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem	27/01/2022	27/01/2022	
									A070504	Premature Discharge of Battery	27/01/2022	27/01/2022	
							F08	Hospitalization or Prolonged Hospitalization	A0705	Battery Problem	27/01/2022	27/01/2022	
									A070504	Premature Discharge of Battery	27/01/2022	27/01/2022	
							F1905	Device Revision or Replacement	A0705	Battery Problem	27/01/2022	27/01/2022	
									A070504	Premature Discharge of Battery	27/01/2022	27/01/2022	
7369	2021	All other reportable incidents	46717	Pacing/defibrillation lead stylet	CT1684	Cardiac pacemakers and associated devices	E200801	Device Embedded In Tissue or Plaque	A0401	Break			
							E2403	No Clinical Signs, Symptoms or Conditions	A0401	Break			
							F15	Recognized Device or Procedural Complication	A0401	Break			
7370	2021	All other reportable incidents	47263	Cardiac resynchronization therapy implantable pacemaker	CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070501	Battery Problem: High Impedance	19/11/2021	19/11/2021	
									A0908	Incorrect, Inadequate or Imprecise Result or Readings	19/11/2021	19/11/2021	
							F08	Hospitalization or Prolonged Hospitalization	A070501	Battery Problem: High Impedance	19/11/2021	19/11/2021	
									A0908	Incorrect, Inadequate or Imprecise Result or Readings	19/11/2021	19/11/2021	
							F1905	Device Revision or Replacement	A070501	Battery Problem: High Impedance	19/11/2021	19/11/2021	
									A0908	Incorrect, Inadequate or Imprecise Result or Readings	19/11/2021	19/11/2021	
					CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A070501	Battery Problem: High Impedance	19/11/2021	19/11/2021	
									A0908	Incorrect, Inadequate or Imprecise Result or Readings	19/11/2021	19/11/2021	
							F08	Hospitalization or Prolonged Hospitalization	A070501	Battery Problem: High Impedance	19/11/2021	19/11/2021	
									A0908	Incorrect, Inadequate or Imprecise Result or Readings	19/11/2021	19/11/2021	
							F1905	Device Revision or Replacement	A070501	Battery Problem: High Impedance	19/11/2021	19/11/2021	
									A0908	Incorrect, Inadequate or Imprecise Result or Readings	19/11/2021	19/11/2021	
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A070501	Battery Problem: High Impedance	19/11/2021	19/11/2021	
									A0908	Incorrect, Inadequate or Imprecise Result or Readings	19/11/2021	19/11/2021	
							F08	Hospitalization or Prolonged Hospitalization	A070501	Battery Problem: High Impedance	19/11/2021	19/11/2021	
									A0908	Incorrect, Inadequate or Imprecise Result or Readings	19/11/2021	19/11/2021	
							F1905	Device Revision or Replacement	A070501	Battery Problem: High Impedance	19/11/2021	19/11/2021	
									A0908	Incorrect, Inadequate or Imprecise Result or Readings	19/11/2021	19/11/2021	
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A070501	Battery Problem: High Impedance	19/11/2021	19/11/2021	
									A0908	Incorrect, Inadequate or Imprecise Result or Readings	19/11/2021	19/11/2021	
							F08	Hospitalization or Prolonged Hospitalization	A070501	Battery Problem: High Impedance	19/11/2021	19/11/2021	
									A0908	Incorrect, Inadequate or Imprecise Result or Readings	19/11/2021	19/11/2021	
							F1905	Device Revision or Replacement	A070501	Battery Problem: High Impedance	19/11/2021	19/11/2021	
									A0908	Incorrect, Inadequate or Imprecise Result or Readings	19/11/2021	19/11/2021	
7371	2022	All other reportable incidents	47265	Dual-chamber implantable pacemaker, rate-responsive	CT141	Electrodes/Leads and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery			
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery			
					CT145	Cardiovascular devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery			
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery			
					CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery			
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery			
					CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery			
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery			
					CT233	Surgical	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery			
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery			
					CT2406	Implantable	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery			
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery			
					CT2407	Active implantable	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery			
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery			
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery			
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery			
					CT2727	Active implant devices/ancillaries	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery			
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery			
					CT2750	Implantable cardiac electrical stimulators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery			
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery			
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery			
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery			
					CT2985	Active	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery			
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery			
					CT3001	Implantable metal	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery			
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery			
					CT3002	Implantable synthetic polymer	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery			
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery			
					CT3006	Implantable inorganic material	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery			
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery			
					CT314	Electrically powered	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery			
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery			
					CT321	Long-term surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery			
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery			
					CT449	Generators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery			
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery			
					CT981	Single use	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery			
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery			
					CT983	Surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery			
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery			
7372	2022	All other reportable incidents	47265	Dual-chamber implantable pacemaker, rate-responsive	CT141	Electrodes/Leads and associated devices	E060101	Asystole	A0712	Pacing Problem	22/07/2022	22/07/2022	
							E060104	Bradycardia	A0712	Pacing Problem	22/07/2022	22/07/2022	

							E060106	Heart Block	A0712	Pacing Problem	22/07/2022	22/07/2022
							E2403	No Clinical Signs, Symptoms or Conditions	A0712	Pacing Problem	22/07/2022	22/07/2022
							F08	Hospitalization or Prolonged Hospitalization	A0712	Pacing Problem	22/07/2022	22/07/2022
							F1203	Life Threatening Illness or Injury	A0712	Pacing Problem	22/07/2022	22/07/2022
							F1905	Device Revision or Replacement	A0712	Pacing Problem	22/07/2022	22/07/2022
					CT145	Cardiovascular devices	E060101	Asystole	A0712	Pacing Problem	22/07/2022	22/07/2022
							E060104	Bradycardia	A0712	Pacing Problem	22/07/2022	22/07/2022
							E060106	Heart Block	A0712	Pacing Problem	22/07/2022	22/07/2022
							E2403	No Clinical Signs, Symptoms or Conditions	A0712	Pacing Problem	22/07/2022	22/07/2022
							F08	Hospitalization or Prolonged Hospitalization	A0712	Pacing Problem	22/07/2022	22/07/2022
							F1203	Life Threatening Illness or Injury	A0712	Pacing Problem	22/07/2022	22/07/2022
							F1905	Device Revision or Replacement	A0712	Pacing Problem	22/07/2022	22/07/2022
					CT1684	Cardiac pacemakers and associated devices	E060101	Asystole	A0712	Pacing Problem	22/07/2022	22/07/2022
							E060104	Bradycardia	A0712	Pacing Problem	22/07/2022	22/07/2022
							E060106	Heart Block	A0712	Pacing Problem	22/07/2022	22/07/2022
							E2403	No Clinical Signs, Symptoms or Conditions	A0712	Pacing Problem	22/07/2022	22/07/2022
							F08	Hospitalization or Prolonged Hospitalization	A0712	Pacing Problem	22/07/2022	22/07/2022
							F1203	Life Threatening Illness or Injury	A0712	Pacing Problem	22/07/2022	22/07/2022
							F1905	Device Revision or Replacement	A0712	Pacing Problem	22/07/2022	22/07/2022
					CT1685	Cardiac pacemakers	E060101	Asystole	A0712	Pacing Problem	22/07/2022	22/07/2022
							E060104	Bradycardia	A0712	Pacing Problem	22/07/2022	22/07/2022
							E060106	Heart Block	A0712	Pacing Problem	22/07/2022	22/07/2022
							E2403	No Clinical Signs, Symptoms or Conditions	A0712	Pacing Problem	22/07/2022	22/07/2022
							F08	Hospitalization or Prolonged Hospitalization	A0712	Pacing Problem	22/07/2022	22/07/2022
							F1203	Life Threatening Illness or Injury	A0712	Pacing Problem	22/07/2022	22/07/2022
							F1905	Device Revision or Replacement	A0712	Pacing Problem	22/07/2022	22/07/2022
					CT233	Surgical	E060101	Asystole	A0712	Pacing Problem	22/07/2022	22/07/2022
							E060104	Bradycardia	A0712	Pacing Problem	22/07/2022	22/07/2022
							E060106	Heart Block	A0712	Pacing Problem	22/07/2022	22/07/2022
							E2403	No Clinical Signs, Symptoms or Conditions	A0712	Pacing Problem	22/07/2022	22/07/2022
							F08	Hospitalization or Prolonged Hospitalization	A0712	Pacing Problem	22/07/2022	22/07/2022
							F1203	Life Threatening Illness or Injury	A0712	Pacing Problem	22/07/2022	22/07/2022
							F1905	Device Revision or Replacement	A0712	Pacing Problem	22/07/2022	22/07/2022
					CT2406	Implantable	E060101	Asystole	A0712	Pacing Problem	22/07/2022	22/07/2022
							E060104	Bradycardia	A0712	Pacing Problem	22/07/2022	22/07/2022
							E060106	Heart Block	A0712	Pacing Problem	22/07/2022	22/07/2022
							E2403	No Clinical Signs, Symptoms or Conditions	A0712	Pacing Problem	22/07/2022	22/07/2022
							F08	Hospitalization or Prolonged Hospitalization	A0712	Pacing Problem	22/07/2022	22/07/2022
							F1203	Life Threatening Illness or Injury	A0712	Pacing Problem	22/07/2022	22/07/2022
							F1905	Device Revision or Replacement	A0712	Pacing Problem	22/07/2022	22/07/2022
					CT2407	Active implantable	E060101	Asystole	A0712	Pacing Problem	22/07/2022	22/07/2022
							E060104	Bradycardia	A0712	Pacing Problem	22/07/2022	22/07/2022
							E060106	Heart Block	A0712	Pacing Problem	22/07/2022	22/07/2022
							E2403	No Clinical Signs, Symptoms or Conditions	A0712	Pacing Problem	22/07/2022	22/07/2022
							F08	Hospitalization or Prolonged Hospitalization	A0712	Pacing Problem	22/07/2022	22/07/2022
							F1203	Life Threatening Illness or Injury	A0712	Pacing Problem	22/07/2022	22/07/2022
							F1905	Device Revision or Replacement	A0712	Pacing Problem	22/07/2022	22/07/2022
					CT257	Implantable cardiac pacemakers	E060101	Asystole	A0712	Pacing Problem	22/07/2022	22/07/2022
							E060104	Bradycardia	A0712	Pacing Problem	22/07/2022	22/07/2022
							E060106	Heart Block	A0712	Pacing Problem	22/07/2022	22/07/2022
							E2403	No Clinical Signs, Symptoms or Conditions	A0712	Pacing Problem	22/07/2022	22/07/2022
							F08	Hospitalization or Prolonged Hospitalization	A0712	Pacing Problem	22/07/2022	22/07/2022
							F1203	Life Threatening Illness or Injury	A0712	Pacing Problem	22/07/2022	22/07/2022
							F1905	Device Revision or Replacement	A0712	Pacing Problem	22/07/2022	22/07/2022
					CT2727	Active implant devices/ancillaries	E060101	Asystole	A0712	Pacing Problem	22/07/2022	22/07/2022
							E060104	Bradycardia	A0712	Pacing Problem	22/07/2022	22/07/2022
							E060106	Heart Block	A0712	Pacing Problem	22/07/2022	22/07/2022
							E2403	No Clinical Signs, Symptoms or Conditions	A0712	Pacing Problem	22/07/2022	22/07/2022
							F08	Hospitalization or Prolonged Hospitalization	A0712	Pacing Problem	22/07/2022	22/07/2022
							F1203	Life Threatening Illness or Injury	A0712	Pacing Problem	22/07/2022	22/07/2022
							F1905	Device Revision or Replacement	A0712	Pacing Problem	22/07/2022	22/07/2022
					CT2750	Implantable cardiac electrical stimulators and associated devices	E060101	Asystole	A0712	Pacing Problem	22/07/2022	22/07/2022
							E060104	Bradycardia	A0712	Pacing Problem	22/07/2022	22/07/2022
							E060106	Heart Block	A0712	Pacing Problem	22/07/2022	22/07/2022
							E2403	No Clinical Signs, Symptoms or Conditions	A0712	Pacing Problem	22/07/2022	22/07/2022
							F08	Hospitalization or Prolonged Hospitalization	A0712	Pacing Problem	22/07/2022	22/07/2022
							F1203	Life Threatening Illness or Injury	A0712	Pacing Problem	22/07/2022	22/07/2022
							F1905	Device Revision or Replacement	A0712	Pacing Problem	22/07/2022	22/07/2022
					CT2866	Implantable cardiac electrical stimulators	E060101	Asystole	A0712	Pacing Problem	22/07/2022	22/07/2022
							E060104	Bradycardia	A0712	Pacing Problem	22/07/2022	22/07/2022
							E060106	Heart Block	A0712	Pacing Problem	22/07/2022	22/07/2022
							E2403	No Clinical Signs, Symptoms or Conditions	A0712	Pacing Problem	22/07/2022	22/07/2022
							F08	Hospitalization or Prolonged Hospitalization	A0712	Pacing Problem	22/07/2022	22/07/2022
							F1203	Life Threatening Illness or Injury	A0712	Pacing Problem	22/07/2022	22/07/2022
							F1905	Device Revision or Replacement	A0712	Pacing Problem	22/07/2022	22/07/2022
					CT2985	Active	E060101	Asystole	A0712	Pacing Problem	22/07/2022	22/07/2022
							E060104	Bradycardia	A0712	Pacing Problem	22/07/2022	22/07/2022
							E060106	Heart Block	A0712	Pacing Problem	22/07/2022	22/07/2022
							E2403	No Clinical Signs, Symptoms or Conditions	A0712	Pacing Problem	22/07/2022	22/07/2022
							F08	Hospitalization or Prolonged Hospitalization	A0712	Pacing Problem	22/07/2022	22/07/2022
							F1203	Life Threatening Illness or Injury	A0712	Pacing Problem	22/07/2022	22/07/2022
							F1905	Device Revision or Replacement	A0712	Pacing Problem	22/07/2022	22/07/2022
					CT3001	Implantable metal	E060101	Asystole	A0712	Pacing Problem	22/07/2022	22/07/2022
							E060104	Bradycardia	A0712	Pacing Problem	22/07/2022	22/07/2022
							E060106	Heart Block	A0712	Pacing Problem	22/07/2022	22/07/2022
							E2403	No Clinical Signs, Symptoms or Conditions	A0712	Pacing Problem	22/07/2022	22/07/2022
							F08	Hospitalization or Prolonged Hospitalization	A0712	Pacing Problem	22/07/2022	22/07/2022
							F1203	Life Threatening Illness or Injury	A0712	Pacing Problem	22/07/2022	22/07/2022

							F1905	Device Revision or Replacement	A0712	Pacing Problem	22/07/2022	22/07/2022
					CT3002	Implantable synthetic polymer	E060101	Asystole	A0712	Pacing Problem	22/07/2022	22/07/2022
							E060104	Bradycardia	A0712	Pacing Problem	22/07/2022	22/07/2022
							E060106	Heart Block	A0712	Pacing Problem	22/07/2022	22/07/2022
							E2403	No Clinical Signs, Symptoms or Conditions	A0712	Pacing Problem	22/07/2022	22/07/2022
							F08	Hospitalization or Prolonged Hospitalization	A0712	Pacing Problem	22/07/2022	22/07/2022
							F1203	Life Threatening Illness or Injury	A0712	Pacing Problem	22/07/2022	22/07/2022
							F1905	Device Revision or Replacement	A0712	Pacing Problem	22/07/2022	22/07/2022
					CT3006	Implantable inorganic material	E060101	Asystole	A0712	Pacing Problem	22/07/2022	22/07/2022
							E060104	Bradycardia	A0712	Pacing Problem	22/07/2022	22/07/2022
							E060106	Heart Block	A0712	Pacing Problem	22/07/2022	22/07/2022
							E2403	No Clinical Signs, Symptoms or Conditions	A0712	Pacing Problem	22/07/2022	22/07/2022
							F08	Hospitalization or Prolonged Hospitalization	A0712	Pacing Problem	22/07/2022	22/07/2022
							F1203	Life Threatening Illness or Injury	A0712	Pacing Problem	22/07/2022	22/07/2022
							F1905	Device Revision or Replacement	A0712	Pacing Problem	22/07/2022	22/07/2022
					CT314	Electrically powered	E060101	Asystole	A0712	Pacing Problem	22/07/2022	22/07/2022
							E060104	Bradycardia	A0712	Pacing Problem	22/07/2022	22/07/2022
							E060106	Heart Block	A0712	Pacing Problem	22/07/2022	22/07/2022
							E2403	No Clinical Signs, Symptoms or Conditions	A0712	Pacing Problem	22/07/2022	22/07/2022
							F08	Hospitalization or Prolonged Hospitalization	A0712	Pacing Problem	22/07/2022	22/07/2022
							F1203	Life Threatening Illness or Injury	A0712	Pacing Problem	22/07/2022	22/07/2022
							F1905	Device Revision or Replacement	A0712	Pacing Problem	22/07/2022	22/07/2022
					CT321	Long-term surgical invasive	E060101	Asystole	A0712	Pacing Problem	22/07/2022	22/07/2022
							E060104	Bradycardia	A0712	Pacing Problem	22/07/2022	22/07/2022
							E060106	Heart Block	A0712	Pacing Problem	22/07/2022	22/07/2022
							E2403	No Clinical Signs, Symptoms or Conditions	A0712	Pacing Problem	22/07/2022	22/07/2022
							F08	Hospitalization or Prolonged Hospitalization	A0712	Pacing Problem	22/07/2022	22/07/2022
							F1203	Life Threatening Illness or Injury	A0712	Pacing Problem	22/07/2022	22/07/2022
							F1905	Device Revision or Replacement	A0712	Pacing Problem	22/07/2022	22/07/2022
					CT449	Generators and associated devices	E060101	Asystole	A0712	Pacing Problem	22/07/2022	22/07/2022
							E060104	Bradycardia	A0712	Pacing Problem	22/07/2022	22/07/2022
							E060106	Heart Block	A0712	Pacing Problem	22/07/2022	22/07/2022
							E2403	No Clinical Signs, Symptoms or Conditions	A0712	Pacing Problem	22/07/2022	22/07/2022
							F08	Hospitalization or Prolonged Hospitalization	A0712	Pacing Problem	22/07/2022	22/07/2022
							F1203	Life Threatening Illness or Injury	A0712	Pacing Problem	22/07/2022	22/07/2022
							F1905	Device Revision or Replacement	A0712	Pacing Problem	22/07/2022	22/07/2022
					CT981	Single use	E060101	Asystole	A0712	Pacing Problem	22/07/2022	22/07/2022
							E060104	Bradycardia	A0712	Pacing Problem	22/07/2022	22/07/2022
							E060106	Heart Block	A0712	Pacing Problem	22/07/2022	22/07/2022
							E2403	No Clinical Signs, Symptoms or Conditions	A0712	Pacing Problem	22/07/2022	22/07/2022
							F08	Hospitalization or Prolonged Hospitalization	A0712	Pacing Problem	22/07/2022	22/07/2022
							F1203	Life Threatening Illness or Injury	A0712	Pacing Problem	22/07/2022	22/07/2022
							F1905	Device Revision or Replacement	A0712	Pacing Problem	22/07/2022	22/07/2022
					CT983	Surgical invasive	E060101	Asystole	A0712	Pacing Problem	22/07/2022	22/07/2022
							E060104	Bradycardia	A0712	Pacing Problem	22/07/2022	22/07/2022
							E060106	Heart Block	A0712	Pacing Problem	22/07/2022	22/07/2022
							E2403	No Clinical Signs, Symptoms or Conditions	A0712	Pacing Problem	22/07/2022	22/07/2022
							F08	Hospitalization or Prolonged Hospitalization	A0712	Pacing Problem	22/07/2022	22/07/2022
							F1203	Life Threatening Illness or Injury	A0712	Pacing Problem	22/07/2022	22/07/2022
							F1905	Device Revision or Replacement	A0712	Pacing Problem	22/07/2022	22/07/2022
7373	2022	All other reportable incidents	35223	Endocardial/interventricular septal pacing lead	CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070102	High Capture Threshold	02/08/2022	02/08/2022
									A072201	High Impedance	02/08/2022	02/08/2022
							F26	No Health Consequences or Impact	A070102	High Capture Threshold	02/08/2022	02/08/2022
									A072201	High Impedance	02/08/2022	02/08/2022
7374	2021	All other reportable incidents	47263	Cardiac resynchronization therapy implantable pacemaker	CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A0908	Incorrect, Inadequate or Imprecise Result or Readings	08/12/2021	08/12/2021
							F08	Hospitalization or Prolonged Hospitalization	A0908	Incorrect, Inadequate or Imprecise Result or Readings	08/12/2021	08/12/2021
							F1905	Device Revision or Replacement	A0908	Incorrect, Inadequate or Imprecise Result or Readings	08/12/2021	08/12/2021
					CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A0908	Incorrect, Inadequate or Imprecise Result or Readings	08/12/2021	08/12/2021
							F08	Hospitalization or Prolonged Hospitalization	A0908	Incorrect, Inadequate or Imprecise Result or Readings	08/12/2021	08/12/2021
							F1905	Device Revision or Replacement	A0908	Incorrect, Inadequate or Imprecise Result or Readings	08/12/2021	08/12/2021
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A0908	Incorrect, Inadequate or Imprecise Result or Readings	08/12/2021	08/12/2021
							F08	Hospitalization or Prolonged Hospitalization	A0908	Incorrect, Inadequate or Imprecise Result or Readings	08/12/2021	08/12/2021
							F1905	Device Revision or Replacement	A0908	Incorrect, Inadequate or Imprecise Result or Readings	08/12/2021	08/12/2021
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A0908	Incorrect, Inadequate or Imprecise Result or Readings	08/12/2021	08/12/2021
							F08	Hospitalization or Prolonged Hospitalization	A0908	Incorrect, Inadequate or Imprecise Result or Readings	08/12/2021	08/12/2021
							F1905	Device Revision or Replacement	A0908	Incorrect, Inadequate or Imprecise Result or Readings	08/12/2021	08/12/2021
7375	2021	All other reportable incidents	47265	Dual-chamber implantable pacemaker, rate-responsive	CT141	Electrodes/Leads and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	15/12/2021	15/12/2021
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	15/12/2021	15/12/2021
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	15/12/2021	15/12/2021
					CT145	Cardiovascular devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	15/12/2021	15/12/2021
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	15/12/2021	15/12/2021
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	15/12/2021	15/12/2021
					CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	15/12/2021	15/12/2021
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	15/12/2021	15/12/2021
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	15/12/2021	15/12/2021
					CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	15/12/2021	15/12/2021
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	15/12/2021	15/12/2021
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	15/12/2021	15/12/2021
					CT233	Surgical	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	15/12/2021	15/12/2021

							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	15/12/2021	15/12/2021
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	15/12/2021	15/12/2021
					CT2406	Implantable	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	15/12/2021	15/12/2021
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	15/12/2021	15/12/2021
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	15/12/2021	15/12/2021
					CT2407	Active implantable	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	15/12/2021	15/12/2021
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	15/12/2021	15/12/2021
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	15/12/2021	15/12/2021
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	15/12/2021	15/12/2021
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	15/12/2021	15/12/2021
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	15/12/2021	15/12/2021
					CT2727	Active implant devices/ancillaries	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	15/12/2021	15/12/2021
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	15/12/2021	15/12/2021
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	15/12/2021	15/12/2021
					CT2750	Implantable cardiac electrical stimulators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	15/12/2021	15/12/2021
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	15/12/2021	15/12/2021
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	15/12/2021	15/12/2021
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	15/12/2021	15/12/2021
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	15/12/2021	15/12/2021
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	15/12/2021	15/12/2021
					CT2985	Active	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	15/12/2021	15/12/2021
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	15/12/2021	15/12/2021
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	15/12/2021	15/12/2021
					CT3001	Implantable metal	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	15/12/2021	15/12/2021
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	15/12/2021	15/12/2021
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	15/12/2021	15/12/2021
					CT3002	Implantable synthetic polymer	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	15/12/2021	15/12/2021
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	15/12/2021	15/12/2021
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	15/12/2021	15/12/2021
					CT3006	Implantable inorganic material	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	15/12/2021	15/12/2021
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	15/12/2021	15/12/2021
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	15/12/2021	15/12/2021
					CT314	Electrically powered	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	15/12/2021	15/12/2021
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	15/12/2021	15/12/2021
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	15/12/2021	15/12/2021
					CT321	Long-term surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	15/12/2021	15/12/2021
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	15/12/2021	15/12/2021
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	15/12/2021	15/12/2021
					CT449	Generators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	15/12/2021	15/12/2021
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	15/12/2021	15/12/2021
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	15/12/2021	15/12/2021
					CT981	Single use	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	15/12/2021	15/12/2021
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	15/12/2021	15/12/2021
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	15/12/2021	15/12/2021
					CT983	Surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	15/12/2021	15/12/2021
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	15/12/2021	15/12/2021
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	15/12/2021	15/12/2021
7376	2021	All other reportable incidents	47265	Dual-chamber implantable pacemaker, rate-responsive	CT141	Electrodes/Leads and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	17/11/2021	17/11/2021
									A0712	Pacing Problem	17/11/2021	17/11/2021
									A1304	Telemetry Discrepancy	17/11/2021	17/11/2021
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	17/11/2021	17/11/2021
									A0712	Pacing Problem	17/11/2021	17/11/2021
									A1304	Telemetry Discrepancy	17/11/2021	17/11/2021
							F1903	Device Explantation	A070504	Premature Discharge of Battery	17/11/2021	17/11/2021
									A0712	Pacing Problem	17/11/2021	17/11/2021
									A1304	Telemetry Discrepancy	17/11/2021	17/11/2021
					CT145	Cardiovascular devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	17/11/2021	17/11/2021
									A0712	Pacing Problem	17/11/2021	17/11/2021
									A1304	Telemetry Discrepancy	17/11/2021	17/11/2021
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	17/11/2021	17/11/2021
									A0712	Pacing Problem	17/11/2021	17/11/2021
									A1304	Telemetry Discrepancy	17/11/2021	17/11/2021
							F1903	Device Explantation	A070504	Premature Discharge of Battery	17/11/2021	17/11/2021
									A0712	Pacing Problem	17/11/2021	17/11/2021
									A1304	Telemetry Discrepancy	17/11/2021	17/11/2021
					CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	17/11/2021	17/11/2021
									A0712	Pacing Problem	17/11/2021	17/11/2021
									A1304	Telemetry Discrepancy	17/11/2021	17/11/2021
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	17/11/2021	17/11/2021
									A0712	Pacing Problem	17/11/2021	17/11/2021
									A1304	Telemetry Discrepancy	17/11/2021	17/11/2021
							F1903	Device Explantation	A070504	Premature Discharge of Battery	17/11/2021	17/11/2021
									A0712	Pacing Problem	17/11/2021	17/11/2021
									A1304	Telemetry Discrepancy	17/11/2021	17/11/2021
					CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	17/11/2021	17/11/2021
									A0712	Pacing Problem	17/11/2021	17/11/2021
									A1304	Telemetry Discrepancy	17/11/2021	17/11/2021
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	17/11/2021	17/11/2021
									A0712	Pacing Problem	17/11/2021	17/11/2021
									A1304	Telemetry Discrepancy	17/11/2021	17/11/2021
							F1903	Device Explantation	A070504	Premature Discharge of Battery	17/11/2021	17/11/2021
									A0712	Pacing Problem	17/11/2021	17/11/2021
									A1304	Telemetry Discrepancy	17/11/2021	17/11/2021
					CT233	Surgical	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	17/11/2021	17/11/2021
									A0712	Pacing Problem	17/11/2021	17/11/2021
									A1304	Telemetry Discrepancy	17/11/2021	17/11/2021
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	17/11/2021	17/11/2021
									A0712	Pacing Problem	17/11/2021	17/11/2021

								A1304	Telemetry Discrepancy	17/11/2021	17/11/2021
						F1903	Device Explantation	A070504	Premature Discharge of Battery	17/11/2021	17/11/2021
								A0712	Pacing Problem	17/11/2021	17/11/2021
								A1304	Telemetry Discrepancy	17/11/2021	17/11/2021
					CT2406	Implantable	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	17/11/2021	17/11/2021
									A0712	Pacing Problem	17/11/2021	17/11/2021
									A1304	Telemetry Discrepancy	17/11/2021	17/11/2021
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	17/11/2021	17/11/2021
									A0712	Pacing Problem	17/11/2021	17/11/2021
									A1304	Telemetry Discrepancy	17/11/2021	17/11/2021
							F1903	Device Explantation	A070504	Premature Discharge of Battery	17/11/2021	17/11/2021
									A0712	Pacing Problem	17/11/2021	17/11/2021
									A1304	Telemetry Discrepancy	17/11/2021	17/11/2021
					CT2407	Active implantable	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	17/11/2021	17/11/2021
									A0712	Pacing Problem	17/11/2021	17/11/2021
									A1304	Telemetry Discrepancy	17/11/2021	17/11/2021
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	17/11/2021	17/11/2021
									A0712	Pacing Problem	17/11/2021	17/11/2021
									A1304	Telemetry Discrepancy	17/11/2021	17/11/2021
							F1903	Device Explantation	A070504	Premature Discharge of Battery	17/11/2021	17/11/2021
									A0712	Pacing Problem	17/11/2021	17/11/2021
									A1304	Telemetry Discrepancy	17/11/2021	17/11/2021
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	17/11/2021	17/11/2021
									A0712	Pacing Problem	17/11/2021	17/11/2021
									A1304	Telemetry Discrepancy	17/11/2021	17/11/2021
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	17/11/2021	17/11/2021
									A0712	Pacing Problem	17/11/2021	17/11/2021
									A1304	Telemetry Discrepancy	17/11/2021	17/11/2021
							F1903	Device Explantation	A070504	Premature Discharge of Battery	17/11/2021	17/11/2021
									A0712	Pacing Problem	17/11/2021	17/11/2021
									A1304	Telemetry Discrepancy	17/11/2021	17/11/2021
					CT2727	Active implant devices/ancillaries	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	17/11/2021	17/11/2021
									A0712	Pacing Problem	17/11/2021	17/11/2021
									A1304	Telemetry Discrepancy	17/11/2021	17/11/2021
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	17/11/2021	17/11/2021
									A0712	Pacing Problem	17/11/2021	17/11/2021
									A1304	Telemetry Discrepancy	17/11/2021	17/11/2021
							F1903	Device Explantation	A070504	Premature Discharge of Battery	17/11/2021	17/11/2021
									A0712	Pacing Problem	17/11/2021	17/11/2021
									A1304	Telemetry Discrepancy	17/11/2021	17/11/2021
					CT2750	Implantable cardiac electrical stimulators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	17/11/2021	17/11/2021
									A0712	Pacing Problem	17/11/2021	17/11/2021
									A1304	Telemetry Discrepancy	17/11/2021	17/11/2021
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	17/11/2021	17/11/2021
									A0712	Pacing Problem	17/11/2021	17/11/2021
									A1304	Telemetry Discrepancy	17/11/2021	17/11/2021
							F1903	Device Explantation	A070504	Premature Discharge of Battery	17/11/2021	17/11/2021
									A0712	Pacing Problem	17/11/2021	17/11/2021
									A1304	Telemetry Discrepancy	17/11/2021	17/11/2021
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	17/11/2021	17/11/2021
									A0712	Pacing Problem	17/11/2021	17/11/2021
									A1304	Telemetry Discrepancy	17/11/2021	17/11/2021
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	17/11/2021	17/11/2021
									A0712	Pacing Problem	17/11/2021	17/11/2021
									A1304	Telemetry Discrepancy	17/11/2021	17/11/2021
							F1903	Device Explantation	A070504	Premature Discharge of Battery	17/11/2021	17/11/2021
									A0712	Pacing Problem	17/11/2021	17/11/2021
									A1304	Telemetry Discrepancy	17/11/2021	17/11/2021
					CT2985	Active	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	17/11/2021	17/11/2021
									A0712	Pacing Problem	17/11/2021	17/11/2021
									A1304	Telemetry Discrepancy	17/11/2021	17/11/2021
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	17/11/2021	17/11/2021
									A0712	Pacing Problem	17/11/2021	17/11/2021
									A1304	Telemetry Discrepancy	17/11/2021	17/11/2021
							F1903	Device Explantation	A070504	Premature Discharge of Battery	17/11/2021	17/11/2021
									A0712	Pacing Problem	17/11/2021	17/11/2021
									A1304	Telemetry Discrepancy	17/11/2021	17/11/2021
					CT3001	Implantable metal	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	17/11/2021	17/11/2021
									A0712	Pacing Problem	17/11/2021	17/11/2021
									A1304	Telemetry Discrepancy	17/11/2021	17/11/2021
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	17/11/2021	17/11/2021
									A0712	Pacing Problem	17/11/2021	17/11/2021
									A1304	Telemetry Discrepancy	17/11/2021	17/11/2021
							F1903	Device Explantation	A070504	Premature Discharge of Battery	17/11/2021	17/11/2021
									A0712	Pacing Problem	17/11/2021	17/11/2021
									A1304	Telemetry Discrepancy	17/11/2021	17/11/2021
					CT3002	Implantable synthetic polymer	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	17/11/2021	17/11/2021
									A0712	Pacing Problem	17/11/2021	17/11/2021
									A1304	Telemetry Discrepancy	17/11/2021	17/11/2021
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	17/11/2021	17/11/2021
									A0712	Pacing Problem	17/11/2021	17/11/2021
									A1304	Telemetry Discrepancy	17/11/2021	17/11/2021
							F1903	Device Explantation	A070504	Premature Discharge of Battery	17/11/2021	17/11/2021
									A0712	Pacing Problem	17/11/2021	17/11/2021
									A1304	Telemetry Discrepancy	17/11/2021	17/11/2021
					CT3006	Implantable inorganic material	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	17/11/2021	17/11/2021
									A0712	Pacing Problem	17/11/2021	17/11/2021
									A1304	Telemetry Discrepancy	17/11/2021	17/11/2021

							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	17/11/2021	17/11/2021
									A0712	Pacing Problem	17/11/2021	17/11/2021
									A1304	Telemetry Discrepancy	17/11/2021	17/11/2021
							F1903	Device Explantation	A070504	Premature Discharge of Battery	17/11/2021	17/11/2021
									A0712	Pacing Problem	17/11/2021	17/11/2021
									A1304	Telemetry Discrepancy	17/11/2021	17/11/2021
					CT314	Electrically powered	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	17/11/2021	17/11/2021
									A0712	Pacing Problem	17/11/2021	17/11/2021
									A1304	Telemetry Discrepancy	17/11/2021	17/11/2021
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	17/11/2021	17/11/2021
									A0712	Pacing Problem	17/11/2021	17/11/2021
									A1304	Telemetry Discrepancy	17/11/2021	17/11/2021
							F1903	Device Explantation	A070504	Premature Discharge of Battery	17/11/2021	17/11/2021
									A0712	Pacing Problem	17/11/2021	17/11/2021
									A1304	Telemetry Discrepancy	17/11/2021	17/11/2021
					CT321	Long-term surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	17/11/2021	17/11/2021
									A0712	Pacing Problem	17/11/2021	17/11/2021
									A1304	Telemetry Discrepancy	17/11/2021	17/11/2021
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	17/11/2021	17/11/2021
									A0712	Pacing Problem	17/11/2021	17/11/2021
									A1304	Telemetry Discrepancy	17/11/2021	17/11/2021
							F1903	Device Explantation	A070504	Premature Discharge of Battery	17/11/2021	17/11/2021
									A0712	Pacing Problem	17/11/2021	17/11/2021
									A1304	Telemetry Discrepancy	17/11/2021	17/11/2021
					CT449	Generators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	17/11/2021	17/11/2021
									A0712	Pacing Problem	17/11/2021	17/11/2021
									A1304	Telemetry Discrepancy	17/11/2021	17/11/2021
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	17/11/2021	17/11/2021
									A0712	Pacing Problem	17/11/2021	17/11/2021
									A1304	Telemetry Discrepancy	17/11/2021	17/11/2021
							F1903	Device Explantation	A070504	Premature Discharge of Battery	17/11/2021	17/11/2021
									A0712	Pacing Problem	17/11/2021	17/11/2021
									A1304	Telemetry Discrepancy	17/11/2021	17/11/2021
					CT981	Single use	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	17/11/2021	17/11/2021
									A0712	Pacing Problem	17/11/2021	17/11/2021
									A1304	Telemetry Discrepancy	17/11/2021	17/11/2021
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	17/11/2021	17/11/2021
									A0712	Pacing Problem	17/11/2021	17/11/2021
									A1304	Telemetry Discrepancy	17/11/2021	17/11/2021
							F1903	Device Explantation	A070504	Premature Discharge of Battery	17/11/2021	17/11/2021
									A0712	Pacing Problem	17/11/2021	17/11/2021
									A1304	Telemetry Discrepancy	17/11/2021	17/11/2021
					CT983	Surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	17/11/2021	17/11/2021
									A0712	Pacing Problem	17/11/2021	17/11/2021
									A1304	Telemetry Discrepancy	17/11/2021	17/11/2021
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	17/11/2021	17/11/2021
									A0712	Pacing Problem	17/11/2021	17/11/2021
									A1304	Telemetry Discrepancy	17/11/2021	17/11/2021
							F1903	Device Explantation	A070504	Premature Discharge of Battery	17/11/2021	17/11/2021
									A0712	Pacing Problem	17/11/2021	17/11/2021
									A1304	Telemetry Discrepancy	17/11/2021	17/11/2021
7377	2021	All other reportable incidents	47263	Cardiac resynchronization therapy implantable pacemaker	CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	03/12/2021	03/12/2021
									A0908	Incorrect, Inadequate or Imprecise Result or Readings	03/12/2021	03/12/2021
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	03/12/2021	03/12/2021
									A0908	Incorrect, Inadequate or Imprecise Result or Readings	03/12/2021	03/12/2021
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	03/12/2021	03/12/2021
									A0908	Incorrect, Inadequate or Imprecise Result or Readings	03/12/2021	03/12/2021
					CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	03/12/2021	03/12/2021
									A0908	Incorrect, Inadequate or Imprecise Result or Readings	03/12/2021	03/12/2021
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	03/12/2021	03/12/2021
									A0908	Incorrect, Inadequate or Imprecise Result or Readings	03/12/2021	03/12/2021
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	03/12/2021	03/12/2021
									A0908	Incorrect, Inadequate or Imprecise Result or Readings	03/12/2021	03/12/2021
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	03/12/2021	03/12/2021
									A0908	Incorrect, Inadequate or Imprecise Result or Readings	03/12/2021	03/12/2021
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	03/12/2021	03/12/2021
									A0908	Incorrect, Inadequate or Imprecise Result or Readings	03/12/2021	03/12/2021
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	03/12/2021	03/12/2021
									A0908	Incorrect, Inadequate or Imprecise Result or Readings	03/12/2021	03/12/2021
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	03/12/2021	03/12/2021
									A0908	Incorrect, Inadequate or Imprecise Result or Readings	03/12/2021	03/12/2021
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	03/12/2021	03/12/2021
									A0908	Incorrect, Inadequate or Imprecise Result or Readings	03/12/2021	03/12/2021
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	03/12/2021	03/12/2021
									A0908	Incorrect, Inadequate or Imprecise Result or Readings	03/12/2021	03/12/2021
7378	2021	All other reportable incidents	47265	Dual-chamber implantable pacemaker, rate-responsive	CT141	Electrodes/Leads and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT145	Cardiovascular devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		

					CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT233	Surgical	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT2406	Implantable	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT2407	Active implantable	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT2727	Active implant devices/ancillaries	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT2750	Implantable cardiac electrical stimulators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT2985	Active	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT3001	Implantable metal	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT3002	Implantable synthetic polymer	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT3006	Implantable inorganic material	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT314	Electrically powered	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT321	Long-term surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT449	Generators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT981	Single use	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT983	Surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
7379	2022	All other reportable incidents	37265	Dual-chamber implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F12	Serious Injury/ Illness/ Impairment	A070504	Premature Discharge of Battery		
							F23	Unexpected Medical Intervention	A070504	Premature Discharge of Battery		
7380	2021	All other reportable incidents	35852	Single-chamber implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E2104	Electric Shock	A070909	Oversensing		
									A071302	Inappropriate/Inadequate Shock/Stimulation		
							E2403	No Clinical Signs, Symptoms or Conditions	A070909	Oversensing		
									A071302	Inappropriate/Inadequate Shock/Stimulation		
							F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A070909	Oversensing		
									A071302	Inappropriate/Inadequate Shock/Stimulation		
7381	2022	All other reportable incidents	47265	Dual-chamber implantable pacemaker, rate-responsive	CT141	Electrodes/Leads and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem	29/09/2022	29/09/2022
									A070504	Premature Discharge of Battery	29/09/2022	29/09/2022
							F08	Hospitalization or Prolonged Hospitalization	A0705	Battery Problem	29/09/2022	29/09/2022
									A070504	Premature Discharge of Battery	29/09/2022	29/09/2022
							F1905	Device Revision or Replacement	A0705	Battery Problem	29/09/2022	29/09/2022
									A070504	Premature Discharge of Battery	29/09/2022	29/09/2022
					CT145	Cardiovascular devices	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem	29/09/2022	29/09/2022
									A070504	Premature Discharge of Battery	29/09/2022	29/09/2022
							F08	Hospitalization or Prolonged Hospitalization	A0705	Battery Problem	29/09/2022	29/09/2022
									A070504	Premature Discharge of Battery	29/09/2022	29/09/2022
							F1905	Device Revision or Replacement	A0705	Battery Problem	29/09/2022	29/09/2022
									A070504	Premature Discharge of Battery	29/09/2022	29/09/2022
					CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem	29/09/2022	29/09/2022
									A070504	Premature Discharge of Battery	29/09/2022	29/09/2022
							F08	Hospitalization or Prolonged Hospitalization	A0705	Battery Problem	29/09/2022	29/09/2022
									A070504	Premature Discharge of Battery	29/09/2022	29/09/2022
							F1905	Device Revision or Replacement	A0705	Battery Problem	29/09/2022	29/09/2022
									A070504	Premature Discharge of Battery	29/09/2022	29/09/2022
					CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem	29/09/2022	29/09/2022
									A070504	Premature Discharge of Battery	29/09/2022	29/09/2022
							F08	Hospitalization or Prolonged Hospitalization	A0705	Battery Problem	29/09/2022	29/09/2022
									A070504	Premature Discharge of Battery	29/09/2022	29/09/2022
							F1905	Device Revision or Replacement	A0705	Battery Problem	29/09/2022	29/09/2022
									A070504	Premature Discharge of Battery	29/09/2022	29/09/2022
					CT233	Surgical	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem	29/09/2022	29/09/2022
									A070504	Premature Discharge of Battery	29/09/2022	29/09/2022
							F08	Hospitalization or Prolonged Hospitalization	A0705	Battery Problem	29/09/2022	29/09/2022
									A070504	Premature Discharge of Battery	29/09/2022	29/09/2022
							F1905	Device Revision or Replacement	A0705	Battery Problem	29/09/2022	29/09/2022
									A070504	Premature Discharge of Battery	29/09/2022	29/09/2022
					CT2406	Implantable	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem	29/09/2022	29/09/2022
									A070504	Premature Discharge of Battery	29/09/2022	29/09/2022
							F08	Hospitalization or Prolonged Hospitalization	A0705	Battery Problem	29/09/2022	29/09/2022
									A070504	Premature Discharge of Battery	29/09/2022	29/09/2022
							F1905	Device Revision or Replacement	A0705	Battery Problem	29/09/2022	29/09/2022
									A070504	Premature Discharge of Battery	29/09/2022	29/09/2022
					CT2407	Active implantable	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem	29/09/2022	29/09/2022
									A070504	Premature Discharge of Battery	29/09/2022	29/09/2022
							F08	Hospitalization or Prolonged Hospitalization	A0705	Battery Problem	29/09/2022	29/09/2022
									A070504	Premature Discharge of Battery	29/09/2022	29/09/2022
							F1905	Device Revision or Replacement	A0705	Battery Problem	29/09/2022	29/09/2022
									A070504	Premature Discharge of Battery	29/09/2022	29/09/2022
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem	29/09/2022	29/09/2022

								A070504	Premature Discharge of Battery	29/09/2022	29/09/2022
						F08	Hospitalization or Prolonged Hospitalization	A0705	Battery Problem	29/09/2022	29/09/2022
								A070504	Premature Discharge of Battery	29/09/2022	29/09/2022
						F1905	Device Revision or Replacement	A0705	Battery Problem	29/09/2022	29/09/2022
								A070504	Premature Discharge of Battery	29/09/2022	29/09/2022
				CT2727	Active implant devices/ancillaries	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem	29/09/2022	29/09/2022
								A070504	Premature Discharge of Battery	29/09/2022	29/09/2022
						F08	Hospitalization or Prolonged Hospitalization	A0705	Battery Problem	29/09/2022	29/09/2022
								A070504	Premature Discharge of Battery	29/09/2022	29/09/2022
						F1905	Device Revision or Replacement	A0705	Battery Problem	29/09/2022	29/09/2022
								A070504	Premature Discharge of Battery	29/09/2022	29/09/2022
				CT2750	Implantable cardiac electrical stimulators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem	29/09/2022	29/09/2022
								A070504	Premature Discharge of Battery	29/09/2022	29/09/2022
						F08	Hospitalization or Prolonged Hospitalization	A0705	Battery Problem	29/09/2022	29/09/2022
								A070504	Premature Discharge of Battery	29/09/2022	29/09/2022
						F1905	Device Revision or Replacement	A0705	Battery Problem	29/09/2022	29/09/2022
								A070504	Premature Discharge of Battery	29/09/2022	29/09/2022
				CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem	29/09/2022	29/09/2022
								A070504	Premature Discharge of Battery	29/09/2022	29/09/2022
						F08	Hospitalization or Prolonged Hospitalization	A0705	Battery Problem	29/09/2022	29/09/2022
								A070504	Premature Discharge of Battery	29/09/2022	29/09/2022
						F1905	Device Revision or Replacement	A0705	Battery Problem	29/09/2022	29/09/2022
								A070504	Premature Discharge of Battery	29/09/2022	29/09/2022
				CT2985	Active	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem	29/09/2022	29/09/2022
								A070504	Premature Discharge of Battery	29/09/2022	29/09/2022
						F08	Hospitalization or Prolonged Hospitalization	A0705	Battery Problem	29/09/2022	29/09/2022
								A070504	Premature Discharge of Battery	29/09/2022	29/09/2022
						F1905	Device Revision or Replacement	A0705	Battery Problem	29/09/2022	29/09/2022
								A070504	Premature Discharge of Battery	29/09/2022	29/09/2022
				CT3001	Implantable metal	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem	29/09/2022	29/09/2022
								A070504	Premature Discharge of Battery	29/09/2022	29/09/2022
						F08	Hospitalization or Prolonged Hospitalization	A0705	Battery Problem	29/09/2022	29/09/2022
								A070504	Premature Discharge of Battery	29/09/2022	29/09/2022
						F1905	Device Revision or Replacement	A0705	Battery Problem	29/09/2022	29/09/2022
								A070504	Premature Discharge of Battery	29/09/2022	29/09/2022
				CT3002	Implantable synthetic polymer	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem	29/09/2022	29/09/2022
								A070504	Premature Discharge of Battery	29/09/2022	29/09/2022
						F08	Hospitalization or Prolonged Hospitalization	A0705	Battery Problem	29/09/2022	29/09/2022
								A070504	Premature Discharge of Battery	29/09/2022	29/09/2022
						F1905	Device Revision or Replacement	A0705	Battery Problem	29/09/2022	29/09/2022
								A070504	Premature Discharge of Battery	29/09/2022	29/09/2022
				CT3006	Implantable inorganic material	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem	29/09/2022	29/09/2022
								A070504	Premature Discharge of Battery	29/09/2022	29/09/2022
						F08	Hospitalization or Prolonged Hospitalization	A0705	Battery Problem	29/09/2022	29/09/2022
								A070504	Premature Discharge of Battery	29/09/2022	29/09/2022
						F1905	Device Revision or Replacement	A0705	Battery Problem	29/09/2022	29/09/2022
								A070504	Premature Discharge of Battery	29/09/2022	29/09/2022
				CT314	Electrically powered	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem	29/09/2022	29/09/2022
								A070504	Premature Discharge of Battery	29/09/2022	29/09/2022
						F08	Hospitalization or Prolonged Hospitalization	A0705	Battery Problem	29/09/2022	29/09/2022
								A070504	Premature Discharge of Battery	29/09/2022	29/09/2022
						F1905	Device Revision or Replacement	A0705	Battery Problem	29/09/2022	29/09/2022
								A070504	Premature Discharge of Battery	29/09/2022	29/09/2022
				CT321	Long-term surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem	29/09/2022	29/09/2022
								A070504	Premature Discharge of Battery	29/09/2022	29/09/2022
						F08	Hospitalization or Prolonged Hospitalization	A0705	Battery Problem	29/09/2022	29/09/2022
								A070504	Premature Discharge of Battery	29/09/2022	29/09/2022
						F1905	Device Revision or Replacement	A0705	Battery Problem	29/09/2022	29/09/2022
								A070504	Premature Discharge of Battery	29/09/2022	29/09/2022
				CT449	Generators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem	29/09/2022	29/09/2022
								A070504	Premature Discharge of Battery	29/09/2022	29/09/2022
						F08	Hospitalization or Prolonged Hospitalization	A0705	Battery Problem	29/09/2022	29/09/2022
								A070504	Premature Discharge of Battery	29/09/2022	29/09/2022
						F1905	Device Revision or Replacement	A0705	Battery Problem	29/09/2022	29/09/2022
								A070504	Premature Discharge of Battery	29/09/2022	29/09/2022
				CT981	Single use	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem	29/09/2022	29/09/2022
								A070504	Premature Discharge of Battery	29/09/2022	29/09/2022
						F08	Hospitalization or Prolonged Hospitalization	A0705	Battery Problem	29/09/2022	29/09/2022
								A070504	Premature Discharge of Battery	29/09/2022	29/09/2022
						F1905	Device Revision or Replacement	A0705	Battery Problem	29/09/2022	29/09/2022
								A070504	Premature Discharge of Battery	29/09/2022	29/09/2022
				CT983	Surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem	29/09/2022	29/09/2022
								A070504	Premature Discharge of Battery	29/09/2022	29/09/2022
						F08	Hospitalization or Prolonged Hospitalization	A0705	Battery Problem	29/09/2022	29/09/2022
								A070504	Premature Discharge of Battery	29/09/2022	29/09/2022
						F1905	Device Revision or Replacement	A0705	Battery Problem	29/09/2022	29/09/2022
								A070504	Premature Discharge of Battery	29/09/2022	29/09/2022
7382	2022	All other reportable incidents	47265	Dual-chamber implantable pacemaker, rate-responsive	CT141	Electrodes/Leads and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT145	Cardiovascular devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT233	Surgical	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT2406	Implantable	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		

						F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery			
					CT2407	Active implantable	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT2727	Active implant devices/ancillaries	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT2750	Implantable cardiac electrical stimulators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT2985	Active	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT3001	Implantable metal	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT3002	Implantable synthetic polymer	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT3006	Implantable inorganic material	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT314	Electrically powered	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT321	Long-term surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT449	Generators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT981	Single use	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT983	Surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
7383	2022	All other reportable incidents	47270	Cardiac resynchronization therapy implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	02/08/2022	02/08/2022
							F1901	Additional Surgery	A070504	Premature Discharge of Battery	02/08/2022	02/08/2022
7384	2021	All other reportable incidents	47265	Dual-chamber implantable pacemaker, rate-responsive	CT141	Electrodes/Leads and associated devices	E0112	Dizziness	A070909	Oversensing	26/11/2021	26/11/2021
									A071204	No Pacing	26/11/2021	26/11/2021
							E2403	No Clinical Signs, Symptoms or Conditions	A070909	Oversensing	26/11/2021	26/11/2021
									A071204	No Pacing	26/11/2021	26/11/2021
							F1901	Additional Surgery	A070909	Oversensing	26/11/2021	26/11/2021
									A071204	No Pacing	26/11/2021	26/11/2021
					CT145	Cardiovascular devices	E0112	Dizziness	A070909	Oversensing	26/11/2021	26/11/2021
									A071204	No Pacing	26/11/2021	26/11/2021
							E2403	No Clinical Signs, Symptoms or Conditions	A070909	Oversensing	26/11/2021	26/11/2021
									A071204	No Pacing	26/11/2021	26/11/2021
							F1901	Additional Surgery	A070909	Oversensing	26/11/2021	26/11/2021
									A071204	No Pacing	26/11/2021	26/11/2021
					CT1684	Cardiac pacemakers and associated devices	E0112	Dizziness	A070909	Oversensing	26/11/2021	26/11/2021
									A071204	No Pacing	26/11/2021	26/11/2021
							E2403	No Clinical Signs, Symptoms or Conditions	A070909	Oversensing	26/11/2021	26/11/2021
									A071204	No Pacing	26/11/2021	26/11/2021
							F1901	Additional Surgery	A070909	Oversensing	26/11/2021	26/11/2021
									A071204	No Pacing	26/11/2021	26/11/2021
					CT1685	Cardiac pacemakers	E0112	Dizziness	A070909	Oversensing	26/11/2021	26/11/2021
									A071204	No Pacing	26/11/2021	26/11/2021
							E2403	No Clinical Signs, Symptoms or Conditions	A070909	Oversensing	26/11/2021	26/11/2021
									A071204	No Pacing	26/11/2021	26/11/2021
							F1901	Additional Surgery	A070909	Oversensing	26/11/2021	26/11/2021
									A071204	No Pacing	26/11/2021	26/11/2021
					CT233	Surgical	E0112	Dizziness	A070909	Oversensing	26/11/2021	26/11/2021
									A071204	No Pacing	26/11/2021	26/11/2021
							E2403	No Clinical Signs, Symptoms or Conditions	A070909	Oversensing	26/11/2021	26/11/2021
									A071204	No Pacing	26/11/2021	26/11/2021
							F1901	Additional Surgery	A070909	Oversensing	26/11/2021	26/11/2021
									A071204	No Pacing	26/11/2021	26/11/2021
					CT2406	Implantable	E0112	Dizziness	A070909	Oversensing	26/11/2021	26/11/2021
									A071204	No Pacing	26/11/2021	26/11/2021
							E2403	No Clinical Signs, Symptoms or Conditions	A070909	Oversensing	26/11/2021	26/11/2021
									A071204	No Pacing	26/11/2021	26/11/2021
							F1901	Additional Surgery	A070909	Oversensing	26/11/2021	26/11/2021
									A071204	No Pacing	26/11/2021	26/11/2021
					CT2407	Active implantable	E0112	Dizziness	A070909	Oversensing	26/11/2021	26/11/2021
									A071204	No Pacing	26/11/2021	26/11/2021
							E2403	No Clinical Signs, Symptoms or Conditions	A070909	Oversensing	26/11/2021	26/11/2021
									A071204	No Pacing	26/11/2021	26/11/2021
							F1901	Additional Surgery	A070909	Oversensing	26/11/2021	26/11/2021
									A071204	No Pacing	26/11/2021	26/11/2021
					CT257	Implantable cardiac pacemakers	E0112	Dizziness	A070909	Oversensing	26/11/2021	26/11/2021
									A071204	No Pacing	26/11/2021	26/11/2021
							E2403	No Clinical Signs, Symptoms or Conditions	A070909	Oversensing	26/11/2021	26/11/2021
									A071204	No Pacing	26/11/2021	26/11/2021
							F1901	Additional Surgery	A070909	Oversensing	26/11/2021	26/11/2021
									A071204	No Pacing	26/11/2021	26/11/2021
					CT2727	Active implant devices/ancillaries	E0112	Dizziness	A070909	Oversensing	26/11/2021	26/11/2021
									A071204	No Pacing	26/11/2021	26/11/2021
							E2403	No Clinical Signs, Symptoms or Conditions	A070909	Oversensing	26/11/2021	26/11/2021
									A071204	No Pacing	26/11/2021	26/11/2021
							F1901	Additional Surgery	A070909	Oversensing	26/11/2021	26/11/2021
									A071204	No Pacing	26/11/2021	26/11/2021
					CT2750	Implantable cardiac electrical stimulators and associated devices	E0112	Dizziness	A070909	Oversensing	26/11/2021	26/11/2021
									A071204	No Pacing	26/11/2021	26/11/2021
							E2403	No Clinical Signs, Symptoms or Conditions	A070909	Oversensing	26/11/2021	26/11/2021

								A071204	No Pacing	26/11/2021	26/11/2021
						F1901	Additional Surgery	A070909	Oversensing	26/11/2021	26/11/2021
								A071204	No Pacing	26/11/2021	26/11/2021
				CT2866	Implantable cardiac electrical stimulators	E0112	Dizziness	A070909	Oversensing	26/11/2021	26/11/2021
								A071204	No Pacing	26/11/2021	26/11/2021
						E2403	No Clinical Signs, Symptoms or Conditions	A070909	Oversensing	26/11/2021	26/11/2021
								A071204	No Pacing	26/11/2021	26/11/2021
						F1901	Additional Surgery	A070909	Oversensing	26/11/2021	26/11/2021
								A071204	No Pacing	26/11/2021	26/11/2021
				CT2985	Active	E0112	Dizziness	A070909	Oversensing	26/11/2021	26/11/2021
								A071204	No Pacing	26/11/2021	26/11/2021
						E2403	No Clinical Signs, Symptoms or Conditions	A070909	Oversensing	26/11/2021	26/11/2021
								A071204	No Pacing	26/11/2021	26/11/2021
						F1901	Additional Surgery	A070909	Oversensing	26/11/2021	26/11/2021
								A071204	No Pacing	26/11/2021	26/11/2021
				CT3001	Implantable metal	E0112	Dizziness	A070909	Oversensing	26/11/2021	26/11/2021
								A071204	No Pacing	26/11/2021	26/11/2021
						E2403	No Clinical Signs, Symptoms or Conditions	A070909	Oversensing	26/11/2021	26/11/2021
								A071204	No Pacing	26/11/2021	26/11/2021
						F1901	Additional Surgery	A070909	Oversensing	26/11/2021	26/11/2021
								A071204	No Pacing	26/11/2021	26/11/2021
				CT3002	Implantable synthetic polymer	E0112	Dizziness	A070909	Oversensing	26/11/2021	26/11/2021
								A071204	No Pacing	26/11/2021	26/11/2021
						E2403	No Clinical Signs, Symptoms or Conditions	A070909	Oversensing	26/11/2021	26/11/2021
								A071204	No Pacing	26/11/2021	26/11/2021
						F1901	Additional Surgery	A070909	Oversensing	26/11/2021	26/11/2021
								A071204	No Pacing	26/11/2021	26/11/2021
				CT3006	Implantable inorganic material	E0112	Dizziness	A070909	Oversensing	26/11/2021	26/11/2021
								A071204	No Pacing	26/11/2021	26/11/2021
						E2403	No Clinical Signs, Symptoms or Conditions	A070909	Oversensing	26/11/2021	26/11/2021
								A071204	No Pacing	26/11/2021	26/11/2021
						F1901	Additional Surgery	A070909	Oversensing	26/11/2021	26/11/2021
								A071204	No Pacing	26/11/2021	26/11/2021
				CT314	Electrically powered	E0112	Dizziness	A070909	Oversensing	26/11/2021	26/11/2021
								A071204	No Pacing	26/11/2021	26/11/2021
						E2403	No Clinical Signs, Symptoms or Conditions	A070909	Oversensing	26/11/2021	26/11/2021
								A071204	No Pacing	26/11/2021	26/11/2021
						F1901	Additional Surgery	A070909	Oversensing	26/11/2021	26/11/2021
								A071204	No Pacing	26/11/2021	26/11/2021
				CT321	Long-term surgical invasive	E0112	Dizziness	A070909	Oversensing	26/11/2021	26/11/2021
								A071204	No Pacing	26/11/2021	26/11/2021
						E2403	No Clinical Signs, Symptoms or Conditions	A070909	Oversensing	26/11/2021	26/11/2021
								A071204	No Pacing	26/11/2021	26/11/2021
						F1901	Additional Surgery	A070909	Oversensing	26/11/2021	26/11/2021
								A071204	No Pacing	26/11/2021	26/11/2021
				CT449	Generators and associated devices	E0112	Dizziness	A070909	Oversensing	26/11/2021	26/11/2021
								A071204	No Pacing	26/11/2021	26/11/2021
						E2403	No Clinical Signs, Symptoms or Conditions	A070909	Oversensing	26/11/2021	26/11/2021
								A071204	No Pacing	26/11/2021	26/11/2021
						F1901	Additional Surgery	A070909	Oversensing	26/11/2021	26/11/2021
								A071204	No Pacing	26/11/2021	26/11/2021
				CT981	Single use	E0112	Dizziness	A070909	Oversensing	26/11/2021	26/11/2021
								A071204	No Pacing	26/11/2021	26/11/2021
						E2403	No Clinical Signs, Symptoms or Conditions	A070909	Oversensing	26/11/2021	26/11/2021
								A071204	No Pacing	26/11/2021	26/11/2021
						F1901	Additional Surgery	A070909	Oversensing	26/11/2021	26/11/2021
								A071204	No Pacing	26/11/2021	26/11/2021
				CT983	Surgical invasive	E0112	Dizziness	A070909	Oversensing	26/11/2021	26/11/2021
								A071204	No Pacing	26/11/2021	26/11/2021
						E2403	No Clinical Signs, Symptoms or Conditions	A070909	Oversensing	26/11/2021	26/11/2021
								A071204	No Pacing	26/11/2021	26/11/2021
						F1901	Additional Surgery	A070909	Oversensing	26/11/2021	26/11/2021
								A071204	No Pacing	26/11/2021	26/11/2021
7385	2022	All other reportable incidents	35224	External pacemaker, epicardial pacing	CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A0709	Device Sensing Problem		
									A2204	Inadequate User Interface		
							F26	No Health Consequences or Impact	A0709	Device Sensing Problem		
									A2204	Inadequate User Interface		
					CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A0709	Device Sensing Problem		
									A2204	Inadequate User Interface		
							F26	No Health Consequences or Impact	A0709	Device Sensing Problem		
									A2204	Inadequate User Interface		
7386	2022	All other reportable incidents	47267	Single-chamber implantable pacemaker, rate-responsive	CT141	Electrodes/Leads and associated devices	E0112	Dizziness	A0712	Pacing Problem		
							E011902	Presyncope	A0712	Pacing Problem		
							E060104	Bradycardia	A0712	Pacing Problem		
							E060106	Heart Block	A0712	Pacing Problem		
							E2403	No Clinical Signs, Symptoms or Conditions	A0712	Pacing Problem		
							F08	Hospitalization or Prolonged Hospitalization	A0712	Pacing Problem		
							F12	Serious Injury/ Illness/ Impairment	A0712	Pacing Problem		
					CT145	Cardiovascular devices	E0112	Dizziness	A0712	Pacing Problem		
							E011902	Presyncope	A0712	Pacing Problem		
							E060104	Bradycardia	A0712	Pacing Problem		
							E060106	Heart Block	A0712	Pacing Problem		
							E2403	No Clinical Signs, Symptoms or Conditions	A0712	Pacing Problem		
							F08	Hospitalization or Prolonged Hospitalization	A0712	Pacing Problem		
							F12	Serious Injury/ Illness/ Impairment	A0712	Pacing Problem		
					CT1684	Cardiac pacemakers and associated devices	E0112	Dizziness	A0712	Pacing Problem		
							E011902	Presyncope	A0712	Pacing Problem		
							E060104	Bradycardia	A0712	Pacing Problem		

							E060106	Heart Block	A0712	Pacing Problem		
							E2403	No Clinical Signs, Symptoms or Conditions	A0712	Pacing Problem		
							F08	Hospitalization or Prolonged Hospitalization	A0712	Pacing Problem		
							F12	Serious Injury/ Illness/ Impairment	A0712	Pacing Problem		
						CT1685	Cardiac pacemakers	E0112	Dizziness	A0712	Pacing Problem		
								E011902	Presyncope	A0712	Pacing Problem		
								E060104	Bradycardia	A0712	Pacing Problem		
								E060106	Heart Block	A0712	Pacing Problem		
								E2403	No Clinical Signs, Symptoms or Conditions	A0712	Pacing Problem		
								F08	Hospitalization or Prolonged Hospitalization	A0712	Pacing Problem		
								F12	Serious Injury/ Illness/ Impairment	A0712	Pacing Problem		
						CT233	Surgical	E0112	Dizziness	A0712	Pacing Problem		
								E011902	Presyncope	A0712	Pacing Problem		
								E060104	Bradycardia	A0712	Pacing Problem		
								E060106	Heart Block	A0712	Pacing Problem		
								E2403	No Clinical Signs, Symptoms or Conditions	A0712	Pacing Problem		
								F08	Hospitalization or Prolonged Hospitalization	A0712	Pacing Problem		
								F12	Serious Injury/ Illness/ Impairment	A0712	Pacing Problem		
						CT2406	Implantable	E0112	Dizziness	A0712	Pacing Problem		
								E011902	Presyncope	A0712	Pacing Problem		
								E060104	Bradycardia	A0712	Pacing Problem		
								E060106	Heart Block	A0712	Pacing Problem		
								E2403	No Clinical Signs, Symptoms or Conditions	A0712	Pacing Problem		
								F08	Hospitalization or Prolonged Hospitalization	A0712	Pacing Problem		
								F12	Serious Injury/ Illness/ Impairment	A0712	Pacing Problem		
						CT2407	Active implantable	E0112	Dizziness	A0712	Pacing Problem		
								E011902	Presyncope	A0712	Pacing Problem		
								E060104	Bradycardia	A0712	Pacing Problem		
								E060106	Heart Block	A0712	Pacing Problem		
								E2403	No Clinical Signs, Symptoms or Conditions	A0712	Pacing Problem		
								F08	Hospitalization or Prolonged Hospitalization	A0712	Pacing Problem		
								F12	Serious Injury/ Illness/ Impairment	A0712	Pacing Problem		
						CT257	Implantable cardiac pacemakers	E0112	Dizziness	A0712	Pacing Problem		
								E011902	Presyncope	A0712	Pacing Problem		
								E060104	Bradycardia	A0712	Pacing Problem		
								E060106	Heart Block	A0712	Pacing Problem		
								E2403	No Clinical Signs, Symptoms or Conditions	A0712	Pacing Problem		
								F08	Hospitalization or Prolonged Hospitalization	A0712	Pacing Problem		
								F12	Serious Injury/ Illness/ Impairment	A0712	Pacing Problem		
						CT2727	Active implant devices/ancillaries	E0112	Dizziness	A0712	Pacing Problem		
								E011902	Presyncope	A0712	Pacing Problem		
								E060104	Bradycardia	A0712	Pacing Problem		
								E060106	Heart Block	A0712	Pacing Problem		
								E2403	No Clinical Signs, Symptoms or Conditions	A0712	Pacing Problem		
								F08	Hospitalization or Prolonged Hospitalization	A0712	Pacing Problem		
								F12	Serious Injury/ Illness/ Impairment	A0712	Pacing Problem		
						CT2750	Implantable cardiac electrical stimulators and associated devices	E0112	Dizziness	A0712	Pacing Problem		
								E011902	Presyncope	A0712	Pacing Problem		
								E060104	Bradycardia	A0712	Pacing Problem		
								E060106	Heart Block	A0712	Pacing Problem		
								E2403	No Clinical Signs, Symptoms or Conditions	A0712	Pacing Problem		
								F08	Hospitalization or Prolonged Hospitalization	A0712	Pacing Problem		
								F12	Serious Injury/ Illness/ Impairment	A0712	Pacing Problem		
						CT2866	Implantable cardiac electrical stimulators	E0112	Dizziness	A0712	Pacing Problem		
								E011902	Presyncope	A0712	Pacing Problem		
								E060104	Bradycardia	A0712	Pacing Problem		
								E060106	Heart Block	A0712	Pacing Problem		
								E2403	No Clinical Signs, Symptoms or Conditions	A0712	Pacing Problem		
								F08	Hospitalization or Prolonged Hospitalization	A0712	Pacing Problem		
								F12	Serious Injury/ Illness/ Impairment	A0712	Pacing Problem		
						CT2985	Active	E0112	Dizziness	A0712	Pacing Problem		
								E011902	Presyncope	A0712	Pacing Problem		
								E060104	Bradycardia	A0712	Pacing Problem		
								E060106	Heart Block	A0712	Pacing Problem		
								E2403	No Clinical Signs, Symptoms or Conditions	A0712	Pacing Problem		
								F08	Hospitalization or Prolonged Hospitalization	A0712	Pacing Problem		
								F12	Serious Injury/ Illness/ Impairment	A0712	Pacing Problem		
						CT3001	Implantable metal	E0112	Dizziness	A0712	Pacing Problem		
								E011902	Presyncope	A0712	Pacing Problem		
								E060104	Bradycardia	A0712	Pacing Problem		
								E060106	Heart Block	A0712	Pacing Problem		
								E2403	No Clinical Signs, Symptoms or Conditions	A0712	Pacing Problem		
								F08	Hospitalization or Prolonged Hospitalization	A0712	Pacing Problem		
								F12	Serious Injury/ Illness/ Impairment	A0712	Pacing Problem		
						CT3002	Implantable synthetic polymer	E0112	Dizziness	A0712	Pacing Problem		
								E011902	Presyncope	A0712	Pacing Problem		
								E060104	Bradycardia	A0712	Pacing Problem		
								E060106	Heart Block	A0712	Pacing Problem		
								E2403	No Clinical Signs, Symptoms or Conditions	A0712	Pacing Problem		
								F08	Hospitalization or Prolonged Hospitalization	A0712	Pacing Problem		
								F12	Serious Injury/ Illness/ Impairment	A0712	Pacing Problem		
						CT3006	Implantable inorganic material	E0112	Dizziness	A0712	Pacing Problem		
								E011902	Presyncope	A0712	Pacing Problem		
								E060104	Bradycardia	A0712	Pacing Problem		
								E060106	Heart Block	A0712	Pacing Problem		
								E2403	No Clinical Signs, Symptoms or Conditions	A0712	Pacing Problem		
								F08	Hospitalization or Prolonged Hospitalization	A0712	Pacing Problem		
								F12	Serious Injury/ Illness/ Impairment	A0712	Pacing Problem		

					CT314	Electrically powered	E0112	Dizziness	A0712	Pacing Problem		
							E011902	Presyncope	A0712	Pacing Problem		
							E060104	Bradycardia	A0712	Pacing Problem		
							E060106	Heart Block	A0712	Pacing Problem		
							E2403	No Clinical Signs, Symptoms or Conditions	A0712	Pacing Problem		
							F08	Hospitalization or Prolonged Hospitalization	A0712	Pacing Problem		
							F12	Serious Injury/ Illness/ Impairment	A0712	Pacing Problem		
					CT321	Long-term surgical invasive	E0112	Dizziness	A0712	Pacing Problem		
							E011902	Presyncope	A0712	Pacing Problem		
							E060104	Bradycardia	A0712	Pacing Problem		
							E060106	Heart Block	A0712	Pacing Problem		
							E2403	No Clinical Signs, Symptoms or Conditions	A0712	Pacing Problem		
							F08	Hospitalization or Prolonged Hospitalization	A0712	Pacing Problem		
							F12	Serious Injury/ Illness/ Impairment	A0712	Pacing Problem		
					CT449	Generators and associated devices	E0112	Dizziness	A0712	Pacing Problem		
							E011902	Presyncope	A0712	Pacing Problem		
							E060104	Bradycardia	A0712	Pacing Problem		
							E060106	Heart Block	A0712	Pacing Problem		
							E2403	No Clinical Signs, Symptoms or Conditions	A0712	Pacing Problem		
							F08	Hospitalization or Prolonged Hospitalization	A0712	Pacing Problem		
							F12	Serious Injury/ Illness/ Impairment	A0712	Pacing Problem		
					CT981	Single use	E0112	Dizziness	A0712	Pacing Problem		
							E011902	Presyncope	A0712	Pacing Problem		
							E060104	Bradycardia	A0712	Pacing Problem		
							E060106	Heart Block	A0712	Pacing Problem		
							E2403	No Clinical Signs, Symptoms or Conditions	A0712	Pacing Problem		
							F08	Hospitalization or Prolonged Hospitalization	A0712	Pacing Problem		
							F12	Serious Injury/ Illness/ Impairment	A0712	Pacing Problem		
					CT983	Surgical invasive	E0112	Dizziness	A0712	Pacing Problem		
							E011902	Presyncope	A0712	Pacing Problem		
							E060104	Bradycardia	A0712	Pacing Problem		
							E060106	Heart Block	A0712	Pacing Problem		
							E2403	No Clinical Signs, Symptoms or Conditions	A0712	Pacing Problem		
							F08	Hospitalization or Prolonged Hospitalization	A0712	Pacing Problem		
							F12	Serious Injury/ Illness/ Impairment	A0712	Pacing Problem		
7387	2021	All other reportable incidents	35852	Single-chamber implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
							F24	Insufficient Health Impact Information	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
7388	2021	All other reportable incidents	47267	Single-chamber implantable pacemaker, rate-responsive	CT141	Electrodes/Leads and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT145	Cardiovascular devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT233	Surgical	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT2406	Implantable	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT2407	Active implantable	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT2727	Active implant devices/ancillaries	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT2750	Implantable cardiac electrical stimulators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT2985	Active	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT3001	Implantable metal	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT3002	Implantable synthetic polymer	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT3006	Implantable inorganic material	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT314	Electrically powered	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT321	Long-term surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT449	Generators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT981	Single use	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT983	Surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
7389	2022	All other reportable incidents	47263	Cardiac resynchronization therapy implantable pacemaker	CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A0908	Incorrect, Inadequate or Imprecise Result or Readings	18/07/2022	18/07/2022
							F08	Hospitalization or Prolonged Hospitalization	A0908	Incorrect, Inadequate or Imprecise Result or Readings	18/07/2022	18/07/2022
							F1905	Device Revision or Replacement	A0908	Incorrect, Inadequate or Imprecise Result or Readings	18/07/2022	18/07/2022
					CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A0908	Incorrect, Inadequate or Imprecise Result or Readings	18/07/2022	18/07/2022
							F08	Hospitalization or Prolonged Hospitalization	A0908	Incorrect, Inadequate or Imprecise Result or Readings	18/07/2022	18/07/2022

							F1905	Device Revision or Replacement	A0908	Incorrect, Inadequate or Imprecise Result or Readings	18/07/2022	18/07/2022
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A0908	Incorrect, Inadequate or Imprecise Result or Readings	18/07/2022	18/07/2022
							F08	Hospitalization or Prolonged Hospitalization	A0908	Incorrect, Inadequate or Imprecise Result or Readings	18/07/2022	18/07/2022
							F1905	Device Revision or Replacement	A0908	Incorrect, Inadequate or Imprecise Result or Readings	18/07/2022	18/07/2022
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A0908	Incorrect, Inadequate or Imprecise Result or Readings	18/07/2022	18/07/2022
							F08	Hospitalization or Prolonged Hospitalization	A0908	Incorrect, Inadequate or Imprecise Result or Readings	18/07/2022	18/07/2022
							F1905	Device Revision or Replacement	A0908	Incorrect, Inadequate or Imprecise Result or Readings	18/07/2022	18/07/2022
7390	2022	All other reportable incidents	37265	Dual-chamber implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A072201	High Impedance		
							F26	No Health Consequences or Impact	A072201	High Impedance		
7391	2022	All other reportable incidents	35852	Single-chamber implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A072201	High Impedance		
							F26	No Health Consequences or Impact	A072201	High Impedance		
7392	2022	All other reportable incidents	47205	Cardiac pulse generator programmer	CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A090805	Incorrect Measurement		
							F23	Unexpected Medical Intervention	A090805	Incorrect Measurement		
7393	2021	All other reportable incidents	47263	Cardiac resynchronization therapy implantable pacemaker	CT1684	Cardiac pacemakers and associated devices	E0112	Dizziness	A070504	Premature Discharge of Battery	10/12/2021	10/12/2021
									A0908	Incorrect, Inadequate or Imprecise Result or Readings	10/12/2021	10/12/2021
							E0138	Undesired Nerve Stimulation	A070504	Premature Discharge of Battery	10/12/2021	10/12/2021
									A0908	Incorrect, Inadequate or Imprecise Result or Readings	10/12/2021	10/12/2021
							E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	10/12/2021	10/12/2021
									A0908	Incorrect, Inadequate or Imprecise Result or Readings	10/12/2021	10/12/2021
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	10/12/2021	10/12/2021
									A0908	Incorrect, Inadequate or Imprecise Result or Readings	10/12/2021	10/12/2021
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	10/12/2021	10/12/2021
									A0908	Incorrect, Inadequate or Imprecise Result or Readings	10/12/2021	10/12/2021
					CT1685	Cardiac pacemakers	E0112	Dizziness	A070504	Premature Discharge of Battery	10/12/2021	10/12/2021
									A0908	Incorrect, Inadequate or Imprecise Result or Readings	10/12/2021	10/12/2021
							E0138	Undesired Nerve Stimulation	A070504	Premature Discharge of Battery	10/12/2021	10/12/2021
									A0908	Incorrect, Inadequate or Imprecise Result or Readings	10/12/2021	10/12/2021
							E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	10/12/2021	10/12/2021
									A0908	Incorrect, Inadequate or Imprecise Result or Readings	10/12/2021	10/12/2021
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	10/12/2021	10/12/2021
									A0908	Incorrect, Inadequate or Imprecise Result or Readings	10/12/2021	10/12/2021
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	10/12/2021	10/12/2021
									A0908	Incorrect, Inadequate or Imprecise Result or Readings	10/12/2021	10/12/2021
					CT257	Implantable cardiac pacemakers	E0112	Dizziness	A070504	Premature Discharge of Battery	10/12/2021	10/12/2021
									A0908	Incorrect, Inadequate or Imprecise Result or Readings	10/12/2021	10/12/2021
							E0138	Undesired Nerve Stimulation	A070504	Premature Discharge of Battery	10/12/2021	10/12/2021
									A0908	Incorrect, Inadequate or Imprecise Result or Readings	10/12/2021	10/12/2021
							E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	10/12/2021	10/12/2021
									A0908	Incorrect, Inadequate or Imprecise Result or Readings	10/12/2021	10/12/2021
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	10/12/2021	10/12/2021
									A0908	Incorrect, Inadequate or Imprecise Result or Readings	10/12/2021	10/12/2021
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	10/12/2021	10/12/2021
									A0908	Incorrect, Inadequate or Imprecise Result or Readings	10/12/2021	10/12/2021
					CT2866	Implantable cardiac electrical stimulators	E0112	Dizziness	A070504	Premature Discharge of Battery	10/12/2021	10/12/2021
									A0908	Incorrect, Inadequate or Imprecise Result or Readings	10/12/2021	10/12/2021
							E0138	Undesired Nerve Stimulation	A070504	Premature Discharge of Battery	10/12/2021	10/12/2021
									A0908	Incorrect, Inadequate or Imprecise Result or Readings	10/12/2021	10/12/2021
							E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	10/12/2021	10/12/2021
									A0908	Incorrect, Inadequate or Imprecise Result or Readings	10/12/2021	10/12/2021
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	10/12/2021	10/12/2021
									A0908	Incorrect, Inadequate or Imprecise Result or Readings	10/12/2021	10/12/2021
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	10/12/2021	10/12/2021
									A0908	Incorrect, Inadequate or Imprecise Result or Readings	10/12/2021	10/12/2021
7394	2021	All other reportable incidents	47270	Cardiac resynchronization therapy implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A072201	High Impedance		
									A1601	Device Alarm System		
							F26	No Health Consequences or Impact	A072201	High Impedance		
									A1601	Device Alarm System		
7395	2021	All other reportable incidents	36241	Pacemaker/defibrillator lead	CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A072201	High Impedance		
							F26	No Health Consequences or Impact	A072201	High Impedance		
7396	2021	All other reportable incidents	17882	Physiologic-monitoring defibrillation system	CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070803	Failure to Power Up		
							F27	Problem Identified During Non-Clinical Procedure	A070803	Failure to Power Up		
7397	2021	All other reportable incidents	35852	Single-chamber implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A160602	Premature End-of-Life Indicator		
							F1903	Device Explantation	A160602	Premature End-of-Life Indicator		
7398	2021	All other reportable incidents	17882	Physiologic-monitoring defibrillation system	CT1684	Cardiac pacemakers and associated devices	E0602	Cardiac Arrest	A070908	Failure to Sense		
							E2403	No Clinical Signs, Symptoms or Conditions	A070908	Failure to Sense		
							F26	No Health Consequences or Impact	A070908	Failure to Sense		
7399	2021	All other reportable incidents	17882	Physiologic-monitoring defibrillation system	CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070803	Failure to Power Up		

							F27	Problem Identified During Non-Clinical Procedure	A070803	Failure to Power Up		
7400	2021	All other reportable incidents	17882	Physiologic-monitoring defibrillation system	CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070803	Failure to Power Up		
							F27	Problem Identified During Non-Clinical Procedure	A070803	Failure to Power Up		
7401	2022	All other reportable incidents	63188	Left-ventricular wireless pacing system sensor/transmitter	CT141	Electrodes/Leads and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070801	Complete Loss of Power		
							F1001	Absence of Treatment	A070801	Complete Loss of Power		
					CT145	Cardiovascular devices	E2403	No Clinical Signs, Symptoms or Conditions	A070801	Complete Loss of Power		
							F1001	Absence of Treatment	A070801	Complete Loss of Power		
					CT1577	Wireless monitoring system transmitters	E2403	No Clinical Signs, Symptoms or Conditions	A070801	Complete Loss of Power		
							F1001	Absence of Treatment	A070801	Complete Loss of Power		
					CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070801	Complete Loss of Power		
							F1001	Absence of Treatment	A070801	Complete Loss of Power		
					CT2317	Transmitters/Receivers/Transceivers	E2403	No Clinical Signs, Symptoms or Conditions	A070801	Complete Loss of Power		
							F1001	Absence of Treatment	A070801	Complete Loss of Power		
					CT233	Surgical	E2403	No Clinical Signs, Symptoms or Conditions	A070801	Complete Loss of Power		
							F1001	Absence of Treatment	A070801	Complete Loss of Power		
					CT2406	Implantable	E2403	No Clinical Signs, Symptoms or Conditions	A070801	Complete Loss of Power		
							F1001	Absence of Treatment	A070801	Complete Loss of Power		
					CT2407	Active implantable	E2403	No Clinical Signs, Symptoms or Conditions	A070801	Complete Loss of Power		
							F1001	Absence of Treatment	A070801	Complete Loss of Power		
					CT255	Transducers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070801	Complete Loss of Power		
							F1001	Absence of Treatment	A070801	Complete Loss of Power		
					CT2726	Physiological monitoring/measuring devices	E2403	No Clinical Signs, Symptoms or Conditions	A070801	Complete Loss of Power		
							F1001	Absence of Treatment	A070801	Complete Loss of Power		
					CT2727	Active implant devices/ancillaries	E2403	No Clinical Signs, Symptoms or Conditions	A070801	Complete Loss of Power		
							F1001	Absence of Treatment	A070801	Complete Loss of Power		
					CT2750	Implantable cardiac electrical stimulators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070801	Complete Loss of Power		
							F1001	Absence of Treatment	A070801	Complete Loss of Power		
					CT2885	Physiological monitoring/measuring system transmitters/receivers/transceivers	E2403	No Clinical Signs, Symptoms or Conditions	A070801	Complete Loss of Power		
							F1001	Absence of Treatment	A070801	Complete Loss of Power		
					CT2985	Active	E2403	No Clinical Signs, Symptoms or Conditions	A070801	Complete Loss of Power		
							F1001	Absence of Treatment	A070801	Complete Loss of Power		
					CT3001	Implantable metal	E2403	No Clinical Signs, Symptoms or Conditions	A070801	Complete Loss of Power		
							F1001	Absence of Treatment	A070801	Complete Loss of Power		
					CT3002	Implantable synthetic polymer	E2403	No Clinical Signs, Symptoms or Conditions	A070801	Complete Loss of Power		
							F1001	Absence of Treatment	A070801	Complete Loss of Power		
					CT3006	Implantable inorganic material	E2403	No Clinical Signs, Symptoms or Conditions	A070801	Complete Loss of Power		
							F1001	Absence of Treatment	A070801	Complete Loss of Power		
					CT314	Electrically powered	E2403	No Clinical Signs, Symptoms or Conditions	A070801	Complete Loss of Power		
							F1001	Absence of Treatment	A070801	Complete Loss of Power		
					CT321	Long-term surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A070801	Complete Loss of Power		
							F1001	Absence of Treatment	A070801	Complete Loss of Power		
					CT384	Sensors and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070801	Complete Loss of Power		
							F1001	Absence of Treatment	A070801	Complete Loss of Power		
					CT449	Generators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070801	Complete Loss of Power		
							F1001	Absence of Treatment	A070801	Complete Loss of Power		
					CT981	Single use	E2403	No Clinical Signs, Symptoms or Conditions	A070801	Complete Loss of Power		
							F1001	Absence of Treatment	A070801	Complete Loss of Power		
					CT983	Surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A070801	Complete Loss of Power		
							F1001	Absence of Treatment	A070801	Complete Loss of Power		
					CT994	Sensors	E2403	No Clinical Signs, Symptoms or Conditions	A070801	Complete Loss of Power		
							F1001	Absence of Treatment	A070801	Complete Loss of Power		
7402	2021	All other reportable incidents	37265	Dual-chamber implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A072201	High Impedance		
							F26	No Health Consequences or Impact	A072201	High Impedance		
7403	2021	All other reportable incidents	35223	Endocardial/interventricular septal pacing lead	CT1684	Cardiac pacemakers and associated devices	E0717	Dyspnea	A070102	High Capture Threshold		
									A070901	Decreased Sensitivity		
									A072202	Low Impedance		
							E2403	No Clinical Signs, Symptoms or Conditions	A070102	High Capture Threshold		
									A070901	Decreased Sensitivity		
									A072202	Low Impedance		
							F12	Serious Injury/ Illness/ Impairment	A070102	High Capture Threshold		
									A070901	Decreased Sensitivity		
									A072202	Low Impedance		
7404	2022	All other reportable incidents	47265	Dual-chamber implantable pacemaker, rate-responsive	CT141	Electrodes/Leads and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem	19/08/2022	19/08/2022
									A070504	Premature Discharge of Battery	19/08/2022	19/08/2022
							F08	Hospitalization or Prolonged Hospitalization	A0705	Battery Problem	19/08/2022	19/08/2022
									A070504	Premature Discharge of Battery	19/08/2022	19/08/2022
							F1905	Device Revision or Replacement	A0705	Battery Problem	19/08/2022	19/08/2022
									A070504	Premature Discharge of Battery	19/08/2022	19/08/2022
					CT145	Cardiovascular devices	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem	19/08/2022	19/08/2022
									A070504	Premature Discharge of Battery	19/08/2022	19/08/2022
							F08	Hospitalization or Prolonged Hospitalization	A0705	Battery Problem	19/08/2022	19/08/2022
									A070504	Premature Discharge of Battery	19/08/2022	19/08/2022
							F1905	Device Revision or Replacement	A0705	Battery Problem	19/08/2022	19/08/2022
									A070504	Premature Discharge of Battery	19/08/2022	19/08/2022
					CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem	19/08/2022	19/08/2022
									A070504	Premature Discharge of Battery	19/08/2022	19/08/2022
							F08	Hospitalization or Prolonged Hospitalization	A0705	Battery Problem	19/08/2022	19/08/2022
									A070504	Premature Discharge of Battery	19/08/2022	19/08/2022
							F1905	Device Revision or Replacement	A0705	Battery Problem	19/08/2022	19/08/2022
									A070504	Premature Discharge of Battery	19/08/2022	19/08/2022
					CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem	19/08/2022	19/08/2022
									A070504	Premature Discharge of Battery	19/08/2022	19/08/2022
							F08	Hospitalization or Prolonged Hospitalization	A0705	Battery Problem	19/08/2022	19/08/2022
									A070504	Premature Discharge of Battery	19/08/2022	19/08/2022
							F1905	Device Revision or Replacement	A0705	Battery Problem	19/08/2022	19/08/2022
									A070504	Premature Discharge of Battery	19/08/2022	19/08/2022
					CT233	Surgical	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem	19/08/2022	19/08/2022
									A070504	Premature Discharge of Battery	19/08/2022	19/08/2022

		F08	Hospitalization or Prolonged Hospitalization	A0705	Battery Problem	19/08/2022	19/08/2022
				A070504	Premature Discharge of Battery	19/08/2022	19/08/2022
		F1905	Device Revision or Replacement	A0705	Battery Problem	19/08/2022	19/08/2022
				A070504	Premature Discharge of Battery	19/08/2022	19/08/2022
CT2406	Implantable	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem	19/08/2022	19/08/2022
				A070504	Premature Discharge of Battery	19/08/2022	19/08/2022
		F08	Hospitalization or Prolonged Hospitalization	A0705	Battery Problem	19/08/2022	19/08/2022
				A070504	Premature Discharge of Battery	19/08/2022	19/08/2022
		F1905	Device Revision or Replacement	A0705	Battery Problem	19/08/2022	19/08/2022
				A070504	Premature Discharge of Battery	19/08/2022	19/08/2022
CT2407	Active implantable	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem	19/08/2022	19/08/2022
				A070504	Premature Discharge of Battery	19/08/2022	19/08/2022
		F08	Hospitalization or Prolonged Hospitalization	A0705	Battery Problem	19/08/2022	19/08/2022
				A070504	Premature Discharge of Battery	19/08/2022	19/08/2022
		F1905	Device Revision or Replacement	A0705	Battery Problem	19/08/2022	19/08/2022
				A070504	Premature Discharge of Battery	19/08/2022	19/08/2022
CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem	19/08/2022	19/08/2022
				A070504	Premature Discharge of Battery	19/08/2022	19/08/2022
		F08	Hospitalization or Prolonged Hospitalization	A0705	Battery Problem	19/08/2022	19/08/2022
				A070504	Premature Discharge of Battery	19/08/2022	19/08/2022
		F1905	Device Revision or Replacement	A0705	Battery Problem	19/08/2022	19/08/2022
				A070504	Premature Discharge of Battery	19/08/2022	19/08/2022
CT2727	Active implant devices/ancillaries	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem	19/08/2022	19/08/2022
				A070504	Premature Discharge of Battery	19/08/2022	19/08/2022
		F08	Hospitalization or Prolonged Hospitalization	A0705	Battery Problem	19/08/2022	19/08/2022
				A070504	Premature Discharge of Battery	19/08/2022	19/08/2022
		F1905	Device Revision or Replacement	A0705	Battery Problem	19/08/2022	19/08/2022
				A070504	Premature Discharge of Battery	19/08/2022	19/08/2022
CT2750	Implantable cardiac electrical stimulators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem	19/08/2022	19/08/2022
				A070504	Premature Discharge of Battery	19/08/2022	19/08/2022
		F08	Hospitalization or Prolonged Hospitalization	A0705	Battery Problem	19/08/2022	19/08/2022
				A070504	Premature Discharge of Battery	19/08/2022	19/08/2022
		F1905	Device Revision or Replacement	A0705	Battery Problem	19/08/2022	19/08/2022
				A070504	Premature Discharge of Battery	19/08/2022	19/08/2022
CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem	19/08/2022	19/08/2022
				A070504	Premature Discharge of Battery	19/08/2022	19/08/2022
		F08	Hospitalization or Prolonged Hospitalization	A0705	Battery Problem	19/08/2022	19/08/2022
				A070504	Premature Discharge of Battery	19/08/2022	19/08/2022
		F1905	Device Revision or Replacement	A0705	Battery Problem	19/08/2022	19/08/2022
				A070504	Premature Discharge of Battery	19/08/2022	19/08/2022
CT2985	Active	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem	19/08/2022	19/08/2022
				A070504	Premature Discharge of Battery	19/08/2022	19/08/2022
		F08	Hospitalization or Prolonged Hospitalization	A0705	Battery Problem	19/08/2022	19/08/2022
				A070504	Premature Discharge of Battery	19/08/2022	19/08/2022
		F1905	Device Revision or Replacement	A0705	Battery Problem	19/08/2022	19/08/2022
				A070504	Premature Discharge of Battery	19/08/2022	19/08/2022
CT3001	Implantable metal	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem	19/08/2022	19/08/2022
				A070504	Premature Discharge of Battery	19/08/2022	19/08/2022
		F08	Hospitalization or Prolonged Hospitalization	A0705	Battery Problem	19/08/2022	19/08/2022
				A070504	Premature Discharge of Battery	19/08/2022	19/08/2022
		F1905	Device Revision or Replacement	A0705	Battery Problem	19/08/2022	19/08/2022
				A070504	Premature Discharge of Battery	19/08/2022	19/08/2022
CT3002	Implantable synthetic polymer	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem	19/08/2022	19/08/2022
				A070504	Premature Discharge of Battery	19/08/2022	19/08/2022
		F08	Hospitalization or Prolonged Hospitalization	A0705	Battery Problem	19/08/2022	19/08/2022
				A070504	Premature Discharge of Battery	19/08/2022	19/08/2022
		F1905	Device Revision or Replacement	A0705	Battery Problem	19/08/2022	19/08/2022
				A070504	Premature Discharge of Battery	19/08/2022	19/08/2022
CT3006	Implantable inorganic material	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem	19/08/2022	19/08/2022
				A070504	Premature Discharge of Battery	19/08/2022	19/08/2022
		F08	Hospitalization or Prolonged Hospitalization	A0705	Battery Problem	19/08/2022	19/08/2022
				A070504	Premature Discharge of Battery	19/08/2022	19/08/2022
		F1905	Device Revision or Replacement	A0705	Battery Problem	19/08/2022	19/08/2022
				A070504	Premature Discharge of Battery	19/08/2022	19/08/2022
CT314	Electrically powered	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem	19/08/2022	19/08/2022
				A070504	Premature Discharge of Battery	19/08/2022	19/08/2022
		F08	Hospitalization or Prolonged Hospitalization	A0705	Battery Problem	19/08/2022	19/08/2022
				A070504	Premature Discharge of Battery	19/08/2022	19/08/2022
		F1905	Device Revision or Replacement	A0705	Battery Problem	19/08/2022	19/08/2022
				A070504	Premature Discharge of Battery	19/08/2022	19/08/2022
CT321	Long-term surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem	19/08/2022	19/08/2022
				A070504	Premature Discharge of Battery	19/08/2022	19/08/2022
		F08	Hospitalization or Prolonged Hospitalization	A0705	Battery Problem	19/08/2022	19/08/2022
				A070504	Premature Discharge of Battery	19/08/2022	19/08/2022
		F1905	Device Revision or Replacement	A0705	Battery Problem	19/08/2022	19/08/2022
				A070504	Premature Discharge of Battery	19/08/2022	19/08/2022
CT449	Generators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem	19/08/2022	19/08/2022
				A070504	Premature Discharge of Battery	19/08/2022	19/08/2022
		F08	Hospitalization or Prolonged Hospitalization	A0705	Battery Problem	19/08/2022	19/08/2022
				A070504	Premature Discharge of Battery	19/08/2022	19/08/2022
		F1905	Device Revision or Replacement	A0705	Battery Problem	19/08/2022	19/08/2022
				A070504	Premature Discharge of Battery	19/08/2022	19/08/2022
CT981	Single use	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem	19/08/2022	19/08/2022
				A070504	Premature Discharge of Battery	19/08/2022	19/08/2022
		F08	Hospitalization or Prolonged Hospitalization	A0705	Battery Problem	19/08/2022	19/08/2022
				A070504	Premature Discharge of Battery	19/08/2022	19/08/2022
		F1905	Device Revision or Replacement	A0705	Battery Problem	19/08/2022	19/08/2022
				A070504	Premature Discharge of Battery	19/08/2022	19/08/2022

					CT983	Surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem	19/08/2022	19/08/2022
									A070504	Premature Discharge of Battery	19/08/2022	19/08/2022
							F08	Hospitalization or Prolonged Hospitalization	A0705	Battery Problem	19/08/2022	19/08/2022
									A070504	Premature Discharge of Battery	19/08/2022	19/08/2022
							F1905	Device Revision or Replacement	A0705	Battery Problem	19/08/2022	19/08/2022
									A070504	Premature Discharge of Battery	19/08/2022	19/08/2022
7405	2022	All other reportable incidents	47263	Cardiac resynchronization therapy implantable pacemaker	CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
7406	2022	All other reportable incidents	47270	Cardiac resynchronization therapy implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
7407	2021	All other reportable incidents	47265	Dual-chamber implantable pacemaker, rate-responsive	CT141	Electrodes/Leads and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
									A072201	High Impedance		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
									A072201	High Impedance		
					CT145	Cardiovascular devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
									A072201	High Impedance		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
									A072201	High Impedance		
					CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
									A072201	High Impedance		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
									A072201	High Impedance		
					CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
									A072201	High Impedance		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
									A072201	High Impedance		
					CT233	Surgical	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
									A072201	High Impedance		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
									A072201	High Impedance		
					CT2406	Implantable	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
									A072201	High Impedance		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
									A072201	High Impedance		
					CT2407	Active implantable	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
									A072201	High Impedance		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
									A072201	High Impedance		
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
									A072201	High Impedance		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
									A072201	High Impedance		
					CT2727	Active implant devices/ancillaries	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
									A072201	High Impedance		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
									A072201	High Impedance		
					CT2750	Implantable cardiac electrical stimulators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
									A072201	High Impedance		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
									A072201	High Impedance		
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
									A072201	High Impedance		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
									A072201	High Impedance		
					CT2985	Active	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
									A072201	High Impedance		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
									A072201	High Impedance		
					CT3001	Implantable metal	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
									A072201	High Impedance		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
									A072201	High Impedance		
					CT3002	Implantable synthetic polymer	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
									A072201	High Impedance		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
									A072201	High Impedance		
					CT3006	Implantable inorganic material	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
									A072201	High Impedance		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
									A072201	High Impedance		
					CT314	Electrically powered	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
									A072201	High Impedance		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
									A072201	High Impedance		
					CT321	Long-term surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
									A072201	High Impedance		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
									A072201	High Impedance		
					CT449	Generators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
									A072201	High Impedance		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
									A072201	High Impedance		

					CT981	Single use	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
									A072201	High Impedance		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
									A072201	High Impedance		
					CT983	Surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
									A072201	High Impedance		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
									A072201	High Impedance		
7408	2021	All other reportable incidents	47265	Dual-chamber implantable pacemaker, rate-responsive	CT141	Electrodes/Leads and associated devices	E020201	Anxiety	A070504	Premature Discharge of Battery	19/11/2021	19/11/2021
							E020204	Malaise	A070504	Premature Discharge of Battery	19/11/2021	19/11/2021
							E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	19/11/2021	19/11/2021
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	19/11/2021	19/11/2021
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	19/11/2021	19/11/2021
					CT145	Cardiovascular devices	E020201	Anxiety	A070504	Premature Discharge of Battery	19/11/2021	19/11/2021
							E020204	Malaise	A070504	Premature Discharge of Battery	19/11/2021	19/11/2021
							E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	19/11/2021	19/11/2021
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	19/11/2021	19/11/2021
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	19/11/2021	19/11/2021
					CT1684	Cardiac pacemakers and associated devices	E020201	Anxiety	A070504	Premature Discharge of Battery	19/11/2021	19/11/2021
							E020204	Malaise	A070504	Premature Discharge of Battery	19/11/2021	19/11/2021
							E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	19/11/2021	19/11/2021
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	19/11/2021	19/11/2021
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	19/11/2021	19/11/2021
					CT1685	Cardiac pacemakers	E020201	Anxiety	A070504	Premature Discharge of Battery	19/11/2021	19/11/2021
							E020204	Malaise	A070504	Premature Discharge of Battery	19/11/2021	19/11/2021
							E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	19/11/2021	19/11/2021
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	19/11/2021	19/11/2021
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	19/11/2021	19/11/2021
					CT233	Surgical	E020201	Anxiety	A070504	Premature Discharge of Battery	19/11/2021	19/11/2021
							E020204	Malaise	A070504	Premature Discharge of Battery	19/11/2021	19/11/2021
							E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	19/11/2021	19/11/2021
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	19/11/2021	19/11/2021
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	19/11/2021	19/11/2021
					CT2406	Implantable	E020201	Anxiety	A070504	Premature Discharge of Battery	19/11/2021	19/11/2021
							E020204	Malaise	A070504	Premature Discharge of Battery	19/11/2021	19/11/2021
							E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	19/11/2021	19/11/2021
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	19/11/2021	19/11/2021
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	19/11/2021	19/11/2021
					CT2407	Active implantable	E020201	Anxiety	A070504	Premature Discharge of Battery	19/11/2021	19/11/2021
							E020204	Malaise	A070504	Premature Discharge of Battery	19/11/2021	19/11/2021
							E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	19/11/2021	19/11/2021
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	19/11/2021	19/11/2021
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	19/11/2021	19/11/2021
					CT257	Implantable cardiac pacemakers	E020201	Anxiety	A070504	Premature Discharge of Battery	19/11/2021	19/11/2021
							E020204	Malaise	A070504	Premature Discharge of Battery	19/11/2021	19/11/2021
							E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	19/11/2021	19/11/2021
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	19/11/2021	19/11/2021
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	19/11/2021	19/11/2021
					CT2727	Active implant devices/ancillaries	E020201	Anxiety	A070504	Premature Discharge of Battery	19/11/2021	19/11/2021
							E020204	Malaise	A070504	Premature Discharge of Battery	19/11/2021	19/11/2021
							E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	19/11/2021	19/11/2021
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	19/11/2021	19/11/2021
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	19/11/2021	19/11/2021
					CT2750	Implantable cardiac electrical stimulators and associated devices	E020201	Anxiety	A070504	Premature Discharge of Battery	19/11/2021	19/11/2021
							E020204	Malaise	A070504	Premature Discharge of Battery	19/11/2021	19/11/2021
							E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	19/11/2021	19/11/2021
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	19/11/2021	19/11/2021
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	19/11/2021	19/11/2021
					CT2866	Implantable cardiac electrical stimulators	E020201	Anxiety	A070504	Premature Discharge of Battery	19/11/2021	19/11/2021
							E020204	Malaise	A070504	Premature Discharge of Battery	19/11/2021	19/11/2021
							E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	19/11/2021	19/11/2021
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	19/11/2021	19/11/2021
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	19/11/2021	19/11/2021
					CT2985	Active	E020201	Anxiety	A070504	Premature Discharge of Battery	19/11/2021	19/11/2021
							E020204	Malaise	A070504	Premature Discharge of Battery	19/11/2021	19/11/2021
							E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	19/11/2021	19/11/2021
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	19/11/2021	19/11/2021
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	19/11/2021	19/11/2021
					CT3001	Implantable metal	E020201	Anxiety	A070504	Premature Discharge of Battery	19/11/2021	19/11/2021
							E020204	Malaise	A070504	Premature Discharge of Battery	19/11/2021	19/11/2021
							E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	19/11/2021	19/11/2021
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	19/11/2021	19/11/2021
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	19/11/2021	19/11/2021
					CT3002	Implantable synthetic polymer	E020201	Anxiety	A070504	Premature Discharge of Battery	19/11/2021	19/11/2021
							E020204	Malaise	A070504	Premature Discharge of Battery	19/11/2021	19/11/2021
							E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	19/11/2021	19/11/2021
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	19/11/2021	19/11/2021
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	19/11/2021	19/11/2021
					CT3006	Implantable inorganic material	E020201	Anxiety	A070504	Premature Discharge of Battery	19/11/2021	19/11/2021
							E020204	Malaise	A070504	Premature Discharge of Battery	19/11/2021	19/11/2021
							E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	19/11/2021	19/11/2021
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	19/11/2021	19/11/2021
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	19/11/2021	19/11/2021
					CT314	Electrically powered	E020201	Anxiety	A070504	Premature Discharge of Battery	19/11/2021	19/11/2021
							E020204	Malaise	A070504	Premature Discharge of Battery	19/11/2021	19/11/2021
							E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	19/11/2021	19/11/2021
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	19/11/2021	19/11/2021
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	19/11/2021	19/11/2021

					CT321	Long-term surgical invasive	E020201	Anxiety	A070504	Premature Discharge of Battery	19/11/2021	19/11/2021
							E020204	Malaise	A070504	Premature Discharge of Battery	19/11/2021	19/11/2021
							E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	19/11/2021	19/11/2021
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	19/11/2021	19/11/2021
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	19/11/2021	19/11/2021
					CT449	Generators and associated devices	E020201	Anxiety	A070504	Premature Discharge of Battery	19/11/2021	19/11/2021
							E020204	Malaise	A070504	Premature Discharge of Battery	19/11/2021	19/11/2021
							E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	19/11/2021	19/11/2021
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	19/11/2021	19/11/2021
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	19/11/2021	19/11/2021
					CT981	Single use	E020201	Anxiety	A070504	Premature Discharge of Battery	19/11/2021	19/11/2021
							E020204	Malaise	A070504	Premature Discharge of Battery	19/11/2021	19/11/2021
							E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	19/11/2021	19/11/2021
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	19/11/2021	19/11/2021
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	19/11/2021	19/11/2021
					CT983	Surgical invasive	E020201	Anxiety	A070504	Premature Discharge of Battery	19/11/2021	19/11/2021
							E020204	Malaise	A070504	Premature Discharge of Battery	19/11/2021	19/11/2021
							E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	19/11/2021	19/11/2021
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	19/11/2021	19/11/2021
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	19/11/2021	19/11/2021
7409	2021	All other reportable incidents	35852	Single-chamber implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E2104	Electric Shock	A071302	Inappropriate/Inadequate Shock/Stimulation		
									A1301	Failure to Read Input Signal		
							E2403	No Clinical Signs, Symptoms or Conditions	A071302	Inappropriate/Inadequate Shock/Stimulation		
									A1301	Failure to Read Input Signal		
							F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A071302	Inappropriate/Inadequate Shock/Stimulation		
									A1301	Failure to Read Input Signal		
7410	2022	All other reportable incidents	35223	Endocardial/interventricular septal pacing lead	CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A0401	Break	04/08/2022	04/08/2022
							F08	Hospitalization or Prolonged Hospitalization	A0401	Break	04/08/2022	04/08/2022
							F1905	Device Revision or Replacement	A0401	Break	04/08/2022	04/08/2022
7411	2022	All other reportable incidents	47265	Dual-chamber implantable pacemaker, rate-responsive	CT141	Electrodes/Leads and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem	27/08/2022	27/08/2022
									A070504	Premature Discharge of Battery	27/08/2022	27/08/2022
							F08	Hospitalization or Prolonged Hospitalization	A0705	Battery Problem	27/08/2022	27/08/2022
									A070504	Premature Discharge of Battery	27/08/2022	27/08/2022
							F1905	Device Revision or Replacement	A0705	Battery Problem	27/08/2022	27/08/2022
									A070504	Premature Discharge of Battery	27/08/2022	27/08/2022
					CT145	Cardiovascular devices	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem	27/08/2022	27/08/2022
									A070504	Premature Discharge of Battery	27/08/2022	27/08/2022
							F08	Hospitalization or Prolonged Hospitalization	A0705	Battery Problem	27/08/2022	27/08/2022
									A070504	Premature Discharge of Battery	27/08/2022	27/08/2022
							F1905	Device Revision or Replacement	A0705	Battery Problem	27/08/2022	27/08/2022
									A070504	Premature Discharge of Battery	27/08/2022	27/08/2022
					CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem	27/08/2022	27/08/2022
									A070504	Premature Discharge of Battery	27/08/2022	27/08/2022
							F08	Hospitalization or Prolonged Hospitalization	A0705	Battery Problem	27/08/2022	27/08/2022
									A070504	Premature Discharge of Battery	27/08/2022	27/08/2022
							F1905	Device Revision or Replacement	A0705	Battery Problem	27/08/2022	27/08/2022
									A070504	Premature Discharge of Battery	27/08/2022	27/08/2022
					CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem	27/08/2022	27/08/2022
									A070504	Premature Discharge of Battery	27/08/2022	27/08/2022
							F08	Hospitalization or Prolonged Hospitalization	A0705	Battery Problem	27/08/2022	27/08/2022
									A070504	Premature Discharge of Battery	27/08/2022	27/08/2022
							F1905	Device Revision or Replacement	A0705	Battery Problem	27/08/2022	27/08/2022
									A070504	Premature Discharge of Battery	27/08/2022	27/08/2022
					CT233	Surgical	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem	27/08/2022	27/08/2022
									A070504	Premature Discharge of Battery	27/08/2022	27/08/2022
							F08	Hospitalization or Prolonged Hospitalization	A0705	Battery Problem	27/08/2022	27/08/2022
									A070504	Premature Discharge of Battery	27/08/2022	27/08/2022
							F1905	Device Revision or Replacement	A0705	Battery Problem	27/08/2022	27/08/2022
									A070504	Premature Discharge of Battery	27/08/2022	27/08/2022
					CT2406	Implantable	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem	27/08/2022	27/08/2022
									A070504	Premature Discharge of Battery	27/08/2022	27/08/2022
							F08	Hospitalization or Prolonged Hospitalization	A0705	Battery Problem	27/08/2022	27/08/2022
									A070504	Premature Discharge of Battery	27/08/2022	27/08/2022
							F1905	Device Revision or Replacement	A0705	Battery Problem	27/08/2022	27/08/2022
									A070504	Premature Discharge of Battery	27/08/2022	27/08/2022
					CT2407	Active implantable	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem	27/08/2022	27/08/2022
									A070504	Premature Discharge of Battery	27/08/2022	27/08/2022
							F08	Hospitalization or Prolonged Hospitalization	A0705	Battery Problem	27/08/2022	27/08/2022
									A070504	Premature Discharge of Battery	27/08/2022	27/08/2022
							F1905	Device Revision or Replacement	A0705	Battery Problem	27/08/2022	27/08/2022
									A070504	Premature Discharge of Battery	27/08/2022	27/08/2022
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem	27/08/2022	27/08/2022
									A070504	Premature Discharge of Battery	27/08/2022	27/08/2022
							F08	Hospitalization or Prolonged Hospitalization	A0705	Battery Problem	27/08/2022	27/08/2022
									A070504	Premature Discharge of Battery	27/08/2022	27/08/2022
							F1905	Device Revision or Replacement	A0705	Battery Problem	27/08/2022	27/08/2022
									A070504	Premature Discharge of Battery	27/08/2022	27/08/2022
					CT2727	Active implant devices/ancillaries	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem	27/08/2022	27/08/2022
									A070504	Premature Discharge of Battery	27/08/2022	27/08/2022
							F08	Hospitalization or Prolonged Hospitalization	A0705	Battery Problem	27/08/2022	27/08/2022
									A070504	Premature Discharge of Battery	27/08/2022	27/08/2022
							F1905	Device Revision or Replacement	A0705	Battery Problem	27/08/2022	27/08/2022
									A070504	Premature Discharge of Battery	27/08/2022	27/08/2022
					CT2750	Implantable cardiac electrical stimulators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem	27/08/2022	27/08/2022
									A070504	Premature Discharge of Battery	27/08/2022	27/08/2022
							F08	Hospitalization or Prolonged Hospitalization	A0705	Battery Problem	27/08/2022	27/08/2022
									A070504	Premature Discharge of Battery	27/08/2022	27/08/2022
							F1905	Device Revision or Replacement	A0705	Battery Problem	27/08/2022	27/08/2022

									A070504	Premature Discharge of Battery	27/08/2022	27/08/2022
				CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem	27/08/2022	27/08/2022
								A070504	Premature Discharge of Battery	27/08/2022	27/08/2022
						F08	Hospitalization or Prolonged Hospitalization	A0705	Battery Problem	27/08/2022	27/08/2022
								A070504	Premature Discharge of Battery	27/08/2022	27/08/2022
						F1905	Device Revision or Replacement	A0705	Battery Problem	27/08/2022	27/08/2022
								A070504	Premature Discharge of Battery	27/08/2022	27/08/2022
				CT2985	Active	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem	27/08/2022	27/08/2022
								A070504	Premature Discharge of Battery	27/08/2022	27/08/2022
						F08	Hospitalization or Prolonged Hospitalization	A0705	Battery Problem	27/08/2022	27/08/2022
								A070504	Premature Discharge of Battery	27/08/2022	27/08/2022
						F1905	Device Revision or Replacement	A0705	Battery Problem	27/08/2022	27/08/2022
								A070504	Premature Discharge of Battery	27/08/2022	27/08/2022
				CT3001	Implantable metal	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem	27/08/2022	27/08/2022
								A070504	Premature Discharge of Battery	27/08/2022	27/08/2022
						F08	Hospitalization or Prolonged Hospitalization	A0705	Battery Problem	27/08/2022	27/08/2022
								A070504	Premature Discharge of Battery	27/08/2022	27/08/2022
						F1905	Device Revision or Replacement	A0705	Battery Problem	27/08/2022	27/08/2022
								A070504	Premature Discharge of Battery	27/08/2022	27/08/2022
				CT3002	Implantable synthetic polymer	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem	27/08/2022	27/08/2022
								A070504	Premature Discharge of Battery	27/08/2022	27/08/2022
						F08	Hospitalization or Prolonged Hospitalization	A0705	Battery Problem	27/08/2022	27/08/2022
								A070504	Premature Discharge of Battery	27/08/2022	27/08/2022
						F1905	Device Revision or Replacement	A0705	Battery Problem	27/08/2022	27/08/2022
								A070504	Premature Discharge of Battery	27/08/2022	27/08/2022
				CT3006	Implantable inorganic material	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem	27/08/2022	27/08/2022
								A070504	Premature Discharge of Battery	27/08/2022	27/08/2022
						F08	Hospitalization or Prolonged Hospitalization	A0705	Battery Problem	27/08/2022	27/08/2022
								A070504	Premature Discharge of Battery	27/08/2022	27/08/2022
						F1905	Device Revision or Replacement	A0705	Battery Problem	27/08/2022	27/08/2022
								A070504	Premature Discharge of Battery	27/08/2022	27/08/2022
				CT314	Electrically powered	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem	27/08/2022	27/08/2022
								A070504	Premature Discharge of Battery	27/08/2022	27/08/2022
						F08	Hospitalization or Prolonged Hospitalization	A0705	Battery Problem	27/08/2022	27/08/2022
								A070504	Premature Discharge of Battery	27/08/2022	27/08/2022
						F1905	Device Revision or Replacement	A0705	Battery Problem	27/08/2022	27/08/2022
								A070504	Premature Discharge of Battery	27/08/2022	27/08/2022
				CT321	Long-term surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem	27/08/2022	27/08/2022
								A070504	Premature Discharge of Battery	27/08/2022	27/08/2022
						F08	Hospitalization or Prolonged Hospitalization	A0705	Battery Problem	27/08/2022	27/08/2022
								A070504	Premature Discharge of Battery	27/08/2022	27/08/2022
						F1905	Device Revision or Replacement	A0705	Battery Problem	27/08/2022	27/08/2022
								A070504	Premature Discharge of Battery	27/08/2022	27/08/2022
				CT449	Generators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem	27/08/2022	27/08/2022
								A070504	Premature Discharge of Battery	27/08/2022	27/08/2022
						F08	Hospitalization or Prolonged Hospitalization	A0705	Battery Problem	27/08/2022	27/08/2022
								A070504	Premature Discharge of Battery	27/08/2022	27/08/2022
						F1905	Device Revision or Replacement	A0705	Battery Problem	27/08/2022	27/08/2022
								A070504	Premature Discharge of Battery	27/08/2022	27/08/2022
				CT981	Single use	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem	27/08/2022	27/08/2022
								A070504	Premature Discharge of Battery	27/08/2022	27/08/2022
						F08	Hospitalization or Prolonged Hospitalization	A0705	Battery Problem	27/08/2022	27/08/2022
								A070504	Premature Discharge of Battery	27/08/2022	27/08/2022
						F1905	Device Revision or Replacement	A0705	Battery Problem	27/08/2022	27/08/2022
								A070504	Premature Discharge of Battery	27/08/2022	27/08/2022
				CT983	Surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem	27/08/2022	27/08/2022
								A070504	Premature Discharge of Battery	27/08/2022	27/08/2022
						F08	Hospitalization or Prolonged Hospitalization	A0705	Battery Problem	27/08/2022	27/08/2022
								A070504	Premature Discharge of Battery	27/08/2022	27/08/2022
						F1905	Device Revision or Replacement	A0705	Battery Problem	27/08/2022	27/08/2022
								A070504	Premature Discharge of Battery	27/08/2022	27/08/2022
7412	2022	All other reportable incidents	35852	Single-chamber implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem	30/11/2022	30/11/2022
									A070504	Premature Discharge of Battery	30/11/2022	30/11/2022
						F08	Hospitalization or Prolonged Hospitalization	A0705	Battery Problem	30/11/2022	30/11/2022
								A070504	Premature Discharge of Battery	30/11/2022	30/11/2022
						F1905	Device Revision or Replacement	A0705	Battery Problem	30/11/2022	30/11/2022
								A070504	Premature Discharge of Battery	30/11/2022	30/11/2022
7413	2021	All other reportable incidents	47265	Dual-chamber implantable pacemaker, rate-responsive	CT141	Electrodes/Leads and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A1605	Reset Problem		
							F26	No Health Consequences or Impact	A1605	Reset Problem		
					CT145	Cardiovascular devices	E2403	No Clinical Signs, Symptoms or Conditions	A1605	Reset Problem		
							F26	No Health Consequences or Impact	A1605	Reset Problem		
					CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A1605	Reset Problem		
							F26	No Health Consequences or Impact	A1605	Reset Problem		
					CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A1605	Reset Problem		
							F26	No Health Consequences or Impact	A1605	Reset Problem		
					CT233	Surgical	E2403	No Clinical Signs, Symptoms or Conditions	A1605	Reset Problem		
							F26	No Health Consequences or Impact	A1605	Reset Problem		
					CT2406	Implantable	E2403	No Clinical Signs, Symptoms or Conditions	A1605	Reset Problem		
							F26	No Health Consequences or Impact	A1605	Reset Problem		
					CT2407	Active implantable	E2403	No Clinical Signs, Symptoms or Conditions	A1605	Reset Problem		
							F26	No Health Consequences or Impact	A1605	Reset Problem		
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A1605	Reset Problem		
							F26	No Health Consequences or Impact	A1605	Reset Problem		
					CT2727	Active implant devices/ancillaries	E2403	No Clinical Signs, Symptoms or Conditions	A1605	Reset Problem		
							F26	No Health Consequences or Impact	A1605	Reset Problem		
					CT2750	Implantable cardiac electrical stimulators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A1605	Reset Problem		
							F26	No Health Consequences or Impact	A1605	Reset Problem		
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A1605	Reset Problem		

						F26	No Health Consequences or Impact	A1605	Reset Problem			
					CT2985	Active	E2403	No Clinical Signs, Symptoms or Conditions	A1605	Reset Problem		
							F26	No Health Consequences or Impact	A1605	Reset Problem		
					CT3001	Implantable metal	E2403	No Clinical Signs, Symptoms or Conditions	A1605	Reset Problem		
							F26	No Health Consequences or Impact	A1605	Reset Problem		
					CT3002	Implantable synthetic polymer	E2403	No Clinical Signs, Symptoms or Conditions	A1605	Reset Problem		
							F26	No Health Consequences or Impact	A1605	Reset Problem		
					CT3006	Implantable inorganic material	E2403	No Clinical Signs, Symptoms or Conditions	A1605	Reset Problem		
							F26	No Health Consequences or Impact	A1605	Reset Problem		
					CT314	Electrically powered	E2403	No Clinical Signs, Symptoms or Conditions	A1605	Reset Problem		
							F26	No Health Consequences or Impact	A1605	Reset Problem		
					CT321	Long-term surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A1605	Reset Problem		
							F26	No Health Consequences or Impact	A1605	Reset Problem		
					CT449	Generators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A1605	Reset Problem		
							F26	No Health Consequences or Impact	A1605	Reset Problem		
					CT981	Single use	E2403	No Clinical Signs, Symptoms or Conditions	A1605	Reset Problem		
							F26	No Health Consequences or Impact	A1605	Reset Problem		
					CT983	Surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A1605	Reset Problem		
							F26	No Health Consequences or Impact	A1605	Reset Problem		
7414	2022	All other reportable incidents	47270	Cardiac resynchronization therapy implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	01/08/2022	01/08/2022
							F1901	Additional Surgery	A070504	Premature Discharge of Battery	01/08/2022	01/08/2022
7415	2021	All other reportable incidents	17882	Physiologic-monitoring defibrillation system	CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A090206	No Display/Image		
							F26	No Health Consequences or Impact	A090206	No Display/Image		
7416	2022	Death	47265	Dual-chamber implantable pacemaker, rate-responsive	CT141	Electrodes/Leads and associated devices	E0119	Loss of Consciousness	A0705	Battery Problem		
							E060104	Bradycardia	A0705	Battery Problem		
							F02	Death	A0705	Battery Problem		
					CT145	Cardiovascular devices	E0119	Loss of Consciousness	A0705	Battery Problem		
							E060104	Bradycardia	A0705	Battery Problem		
							F02	Death	A0705	Battery Problem		
					CT1684	Cardiac pacemakers and associated devices	E0119	Loss of Consciousness	A0705	Battery Problem		
							E060104	Bradycardia	A0705	Battery Problem		
							F02	Death	A0705	Battery Problem		
					CT1685	Cardiac pacemakers	E0119	Loss of Consciousness	A0705	Battery Problem		
							E060104	Bradycardia	A0705	Battery Problem		
							F02	Death	A0705	Battery Problem		
					CT233	Surgical	E0119	Loss of Consciousness	A0705	Battery Problem		
							E060104	Bradycardia	A0705	Battery Problem		
							F02	Death	A0705	Battery Problem		
					CT2406	Implantable	E0119	Loss of Consciousness	A0705	Battery Problem		
							E060104	Bradycardia	A0705	Battery Problem		
							F02	Death	A0705	Battery Problem		
					CT2407	Active implantable	E0119	Loss of Consciousness	A0705	Battery Problem		
							E060104	Bradycardia	A0705	Battery Problem		
							F02	Death	A0705	Battery Problem		
					CT257	Implantable cardiac pacemakers	E0119	Loss of Consciousness	A0705	Battery Problem		
							E060104	Bradycardia	A0705	Battery Problem		
							F02	Death	A0705	Battery Problem		
					CT2727	Active implant devices/ancillaries	E0119	Loss of Consciousness	A0705	Battery Problem		
							E060104	Bradycardia	A0705	Battery Problem		
							F02	Death	A0705	Battery Problem		
					CT2750	Implantable cardiac electrical stimulators and associated devices	E0119	Loss of Consciousness	A0705	Battery Problem		
							E060104	Bradycardia	A0705	Battery Problem		
							F02	Death	A0705	Battery Problem		
					CT2866	Implantable cardiac electrical stimulators	E0119	Loss of Consciousness	A0705	Battery Problem		
							E060104	Bradycardia	A0705	Battery Problem		
							F02	Death	A0705	Battery Problem		
					CT2985	Active	E0119	Loss of Consciousness	A0705	Battery Problem		
							E060104	Bradycardia	A0705	Battery Problem		
							F02	Death	A0705	Battery Problem		
					CT3001	Implantable metal	E0119	Loss of Consciousness	A0705	Battery Problem		
							E060104	Bradycardia	A0705	Battery Problem		
							F02	Death	A0705	Battery Problem		
					CT3002	Implantable synthetic polymer	E0119	Loss of Consciousness	A0705	Battery Problem		
							E060104	Bradycardia	A0705	Battery Problem		
							F02	Death	A0705	Battery Problem		
					CT3006	Implantable inorganic material	E0119	Loss of Consciousness	A0705	Battery Problem		
							E060104	Bradycardia	A0705	Battery Problem		
							F02	Death	A0705	Battery Problem		
					CT314	Electrically powered	E0119	Loss of Consciousness	A0705	Battery Problem		
							E060104	Bradycardia	A0705	Battery Problem		
							F02	Death	A0705	Battery Problem		
					CT321	Long-term surgical invasive	E0119	Loss of Consciousness	A0705	Battery Problem		
							E060104	Bradycardia	A0705	Battery Problem		
							F02	Death	A0705	Battery Problem		
					CT449	Generators and associated devices	E0119	Loss of Consciousness	A0705	Battery Problem		
							E060104	Bradycardia	A0705	Battery Problem		
							F02	Death	A0705	Battery Problem		
					CT981	Single use	E0119	Loss of Consciousness	A0705	Battery Problem		
							E060104	Bradycardia	A0705	Battery Problem		
							F02	Death	A0705	Battery Problem		
					CT983	Surgical invasive	E0119	Loss of Consciousness	A0705	Battery Problem		
							E060104	Bradycardia	A0705	Battery Problem		
							F02	Death	A0705	Battery Problem		
7417	2022	All other reportable incidents	47267	Single-chamber implantable pacemaker, rate-responsive	CT141	Electrodes/Leads and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A090805	Incorrect Measurement		
							F26	No Health Consequences or Impact	A090805	Incorrect Measurement		
					CT145	Cardiovascular devices	E2403	No Clinical Signs, Symptoms or Conditions	A090805	Incorrect Measurement		
							F26	No Health Consequences or Impact	A090805	Incorrect Measurement		
					CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A090805	Incorrect Measurement		

						F26	No Health Consequences or Impact	A090805	Incorrect Measurement			
				CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A090805	Incorrect Measurement			
						F26	No Health Consequences or Impact	A090805	Incorrect Measurement			
				CT233	Surgical	E2403	No Clinical Signs, Symptoms or Conditions	A090805	Incorrect Measurement			
						F26	No Health Consequences or Impact	A090805	Incorrect Measurement			
				CT2406	Implantable	E2403	No Clinical Signs, Symptoms or Conditions	A090805	Incorrect Measurement			
						F26	No Health Consequences or Impact	A090805	Incorrect Measurement			
				CT2407	Active implantable	E2403	No Clinical Signs, Symptoms or Conditions	A090805	Incorrect Measurement			
						F26	No Health Consequences or Impact	A090805	Incorrect Measurement			
				CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A090805	Incorrect Measurement			
						F26	No Health Consequences or Impact	A090805	Incorrect Measurement			
				CT2727	Active implant devices/ancillaries	E2403	No Clinical Signs, Symptoms or Conditions	A090805	Incorrect Measurement			
						F26	No Health Consequences or Impact	A090805	Incorrect Measurement			
				CT2750	Implantable cardiac electrical stimulators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A090805	Incorrect Measurement			
						F26	No Health Consequences or Impact	A090805	Incorrect Measurement			
				CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A090805	Incorrect Measurement			
						F26	No Health Consequences or Impact	A090805	Incorrect Measurement			
				CT2985	Active	E2403	No Clinical Signs, Symptoms or Conditions	A090805	Incorrect Measurement			
						F26	No Health Consequences or Impact	A090805	Incorrect Measurement			
				CT3001	Implantable metal	E2403	No Clinical Signs, Symptoms or Conditions	A090805	Incorrect Measurement			
						F26	No Health Consequences or Impact	A090805	Incorrect Measurement			
				CT3002	Implantable synthetic polymer	E2403	No Clinical Signs, Symptoms or Conditions	A090805	Incorrect Measurement			
						F26	No Health Consequences or Impact	A090805	Incorrect Measurement			
				CT3006	Implantable inorganic material	E2403	No Clinical Signs, Symptoms or Conditions	A090805	Incorrect Measurement			
						F26	No Health Consequences or Impact	A090805	Incorrect Measurement			
				CT314	Electrically powered	E2403	No Clinical Signs, Symptoms or Conditions	A090805	Incorrect Measurement			
						F26	No Health Consequences or Impact	A090805	Incorrect Measurement			
				CT321	Long-term surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A090805	Incorrect Measurement			
						F26	No Health Consequences or Impact	A090805	Incorrect Measurement			
				CT449	Generators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A090805	Incorrect Measurement			
						F26	No Health Consequences or Impact	A090805	Incorrect Measurement			
				CT981	Single use	E2403	No Clinical Signs, Symptoms or Conditions	A090805	Incorrect Measurement			
						F26	No Health Consequences or Impact	A090805	Incorrect Measurement			
				CT983	Surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A090805	Incorrect Measurement			
						F26	No Health Consequences or Impact	A090805	Incorrect Measurement			
7418	2022	All other reportable incidents	35852	Single-chamber implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	18/07/2022	18/07/2022
						F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	18/07/2022	18/07/2022
						F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	18/07/2022	18/07/2022
7419	2022	All other reportable incidents	47263	Cardiac resynchronization therapy implantable pacemaker	CT1684	Cardiac pacemakers and associated devices	E011903	Syncope/Fainting	A0203	Defective Device	27/07/2022	27/07/2022
									A1107	Data Problem	27/07/2022	27/07/2022
							E060110	Ventricular Fibrillation	A0203	Defective Device	27/07/2022	27/07/2022
									A1107	Data Problem	27/07/2022	27/07/2022
							E0602	Cardiac Arrest	A0203	Defective Device	27/07/2022	27/07/2022
									A1107	Data Problem	27/07/2022	27/07/2022
							E2403	No Clinical Signs, Symptoms or Conditions	A0203	Defective Device	27/07/2022	27/07/2022
									A1107	Data Problem	27/07/2022	27/07/2022
							F08	Hospitalization or Prolonged Hospitalization	A0203	Defective Device	27/07/2022	27/07/2022
									A1107	Data Problem	27/07/2022	27/07/2022
							F15	Recognized Device or Procedural Complication	A0203	Defective Device	27/07/2022	27/07/2022
									A1107	Data Problem	27/07/2022	27/07/2022
							F1905	Device Revision or Replacement	A0203	Defective Device	27/07/2022	27/07/2022
									A1107	Data Problem	27/07/2022	27/07/2022
				CT1685	Cardiac pacemakers	E011903	Syncope/Fainting	A0203	Defective Device	27/07/2022	27/07/2022
									A1107	Data Problem	27/07/2022	27/07/2022
							E060110	Ventricular Fibrillation	A0203	Defective Device	27/07/2022	27/07/2022
									A1107	Data Problem	27/07/2022	27/07/2022
							E0602	Cardiac Arrest	A0203	Defective Device	27/07/2022	27/07/2022
									A1107	Data Problem	27/07/2022	27/07/2022
							E2403	No Clinical Signs, Symptoms or Conditions	A0203	Defective Device	27/07/2022	27/07/2022
									A1107	Data Problem	27/07/2022	27/07/2022
							F08	Hospitalization or Prolonged Hospitalization	A0203	Defective Device	27/07/2022	27/07/2022
									A1107	Data Problem	27/07/2022	27/07/2022
							F15	Recognized Device or Procedural Complication	A0203	Defective Device	27/07/2022	27/07/2022
									A1107	Data Problem	27/07/2022	27/07/2022
							F1905	Device Revision or Replacement	A0203	Defective Device	27/07/2022	27/07/2022
									A1107	Data Problem	27/07/2022	27/07/2022
				CT257	Implantable cardiac pacemakers	E011903	Syncope/Fainting	A0203	Defective Device	27/07/2022	27/07/2022
									A1107	Data Problem	27/07/2022	27/07/2022
							E060110	Ventricular Fibrillation	A0203	Defective Device	27/07/2022	27/07/2022
									A1107	Data Problem	27/07/2022	27/07/2022
							E0602	Cardiac Arrest	A0203	Defective Device	27/07/2022	27/07/2022
									A1107	Data Problem	27/07/2022	27/07/2022
							E2403	No Clinical Signs, Symptoms or Conditions	A0203	Defective Device	27/07/2022	27/07/2022
									A1107	Data Problem	27/07/2022	27/07/2022
							F08	Hospitalization or Prolonged Hospitalization	A0203	Defective Device	27/07/2022	27/07/2022
									A1107	Data Problem	27/07/2022	27/07/2022
							F15	Recognized Device or Procedural Complication	A0203	Defective Device	27/07/2022	27/07/2022
									A1107	Data Problem	27/07/2022	27/07/2022
							F1905	Device Revision or Replacement	A0203	Defective Device	27/07/2022	27/07/2022
									A1107	Data Problem	27/07/2022	27/07/2022
				CT2866	Implantable cardiac electrical stimulators	E011903	Syncope/Fainting	A0203	Defective Device	27/07/2022	27/07/2022
									A1107	Data Problem	27/07/2022	27/07/2022
							E060110	Ventricular Fibrillation	A0203	Defective Device	27/07/2022	27/07/2022
									A1107	Data Problem	27/07/2022	27/07/2022
							E0602	Cardiac Arrest	A0203	Defective Device	27/07/2022	27/07/2022
									A1107	Data Problem	27/07/2022	27/07/2022
							E2403	No Clinical Signs, Symptoms or Conditions	A0203	Defective Device	27/07/2022	27/07/2022
									A1107	Data Problem	27/07/2022	27/07/2022

							F08	Hospitalization or Prolonged Hospitalization	A0203	Defective Device	27/07/2022	27/07/2022
									A1107	Data Problem	27/07/2022	27/07/2022
							F15	Recognized Device or Procedural Complication	A0203	Defective Device	27/07/2022	27/07/2022
									A1107	Data Problem	27/07/2022	27/07/2022
							F1905	Device Revision or Replacement	A0203	Defective Device	27/07/2022	27/07/2022
									A1107	Data Problem	27/07/2022	27/07/2022
7420	2021	All other reportable incidents	35852	Single-chamber implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	11/01/2022	11/01/2022
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	11/01/2022	11/01/2022
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	11/01/2022	11/01/2022
7421	2021	All other reportable incidents	35223	Endocardial/interventricular septal pacing lead	CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A072202	Low Impedance		
							F26	No Health Consequences or Impact	A072202	Low Impedance		
7422	2022	All other reportable incidents	17882	Physiologic-monitoring defibrillation system	CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070908	Failure to Sense		
							F27	Problem Identified During Non-Clinical Procedure	A070908	Failure to Sense		
7423	2022	All other reportable incidents	47263	Cardiac resynchronization therapy implantable pacemaker	CT1684	Cardiac pacemakers and associated devices	E2401	Insufficient Information	A0908	Incorrect, Inadequate or Imprecise Result or Readings	28/07/2022	28/07/2022
							E2403	No Clinical Signs, Symptoms or Conditions	A0908	Incorrect, Inadequate or Imprecise Result or Readings	28/07/2022	28/07/2022
							F08	Hospitalization or Prolonged Hospitalization	A0908	Incorrect, Inadequate or Imprecise Result or Readings	28/07/2022	28/07/2022
							F1905	Device Revision or Replacement	A0908	Incorrect, Inadequate or Imprecise Result or Readings	28/07/2022	28/07/2022
					CT1685	Cardiac pacemakers	E2401	Insufficient Information	A0908	Incorrect, Inadequate or Imprecise Result or Readings	28/07/2022	28/07/2022
							E2403	No Clinical Signs, Symptoms or Conditions	A0908	Incorrect, Inadequate or Imprecise Result or Readings	28/07/2022	28/07/2022
							F08	Hospitalization or Prolonged Hospitalization	A0908	Incorrect, Inadequate or Imprecise Result or Readings	28/07/2022	28/07/2022
							F1905	Device Revision or Replacement	A0908	Incorrect, Inadequate or Imprecise Result or Readings	28/07/2022	28/07/2022
					CT257	Implantable cardiac pacemakers	E2401	Insufficient Information	A0908	Incorrect, Inadequate or Imprecise Result or Readings	28/07/2022	28/07/2022
							E2403	No Clinical Signs, Symptoms or Conditions	A0908	Incorrect, Inadequate or Imprecise Result or Readings	28/07/2022	28/07/2022
							F08	Hospitalization or Prolonged Hospitalization	A0908	Incorrect, Inadequate or Imprecise Result or Readings	28/07/2022	28/07/2022
							F1905	Device Revision or Replacement	A0908	Incorrect, Inadequate or Imprecise Result or Readings	28/07/2022	28/07/2022
					CT2866	Implantable cardiac electrical stimulators	E2401	Insufficient Information	A0908	Incorrect, Inadequate or Imprecise Result or Readings	28/07/2022	28/07/2022
							E2403	No Clinical Signs, Symptoms or Conditions	A0908	Incorrect, Inadequate or Imprecise Result or Readings	28/07/2022	28/07/2022
							F08	Hospitalization or Prolonged Hospitalization	A0908	Incorrect, Inadequate or Imprecise Result or Readings	28/07/2022	28/07/2022
							F1905	Device Revision or Replacement	A0908	Incorrect, Inadequate or Imprecise Result or Readings	28/07/2022	28/07/2022
7424	2021	All other reportable incidents	47205	Cardiac pulse generator programmer	CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A1104	Computer Operating System Problem		
							F23	Unexpected Medical Intervention	A1104	Computer Operating System Problem		
7425	2022	All other reportable incidents	47265	Dual-chamber implantable pacemaker, rate-responsive	CT141	Electrodes/Leads and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT145	Cardiovascular devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT233	Surgical	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT2406	Implantable	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT2407	Active implantable	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT2727	Active implant devices/ancillaries	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT2750	Implantable cardiac electrical stimulators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT2985	Active	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT3001	Implantable metal	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT3002	Implantable synthetic polymer	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT3006	Implantable inorganic material	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT314	Electrically powered	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT321	Long-term surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT449	Generators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT981	Single use	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT983	Surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
7426	2021	All other reportable incidents	36241	Pacemaker/defibrillator lead	CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A150206	Difficult to Insert		
							F1908	Prolonged Surgery	A150206	Difficult to Insert		
7427	2021	All other reportable incidents	47265	Dual-chamber implantable pacemaker, rate-responsive	CT141	Electrodes/Leads and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT145	Cardiovascular devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		

							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT233	Surgical	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT2406	Implantable	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT2407	Active implantable	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT2727	Active implant devices/ancillaries	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT2750	Implantable cardiac electrical stimulators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT2985	Active	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT3001	Implantable metal	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT3002	Implantable synthetic polymer	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT3006	Implantable inorganic material	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT314	Electrically powered	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT321	Long-term surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT449	Generators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT981	Single use	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT983	Surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
7428	2022	All other reportable incidents	47265	Dual-chamber implantable pacemaker, rate-responsive	CT141	Electrodes/Leads and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT145	Cardiovascular devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT233	Surgical	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT2406	Implantable	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT2407	Active implantable	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT2727	Active implant devices/ancillaries	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT2750	Implantable cardiac electrical stimulators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT2985	Active	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT3001	Implantable metal	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT3002	Implantable synthetic polymer	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT3006	Implantable inorganic material	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT314	Electrically powered	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT321	Long-term surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT449	Generators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT981	Single use	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT983	Surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
7429	2022	All other reportable incidents	36241	Pacemaker/defibrillator lead	CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A072201	High Impedance		
							F26	No Health Consequences or Impact	A072201	High Impedance		
7430	2021	All other reportable incidents	63188	Left-ventricular wireless pacing system sensor/transmitter	CT141	Electrodes/Leads and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070103	Intermittent Capture		
					CT145	Cardiovascular devices	E2403	No Clinical Signs, Symptoms or Conditions	A070103	Intermittent Capture		
					CT1577	Wireless monitoring system transmitters	E2403	No Clinical Signs, Symptoms or Conditions	A070103	Intermittent Capture		
					CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070103	Intermittent Capture		
					CT2317	Transmitters/Receivers/Transceivers	E2403	No Clinical Signs, Symptoms or Conditions	A070103	Intermittent Capture		
					CT233	Surgical	E2403	No Clinical Signs, Symptoms or Conditions	A070103	Intermittent Capture		
					CT2406	Implantable	E2403	No Clinical Signs, Symptoms or Conditions	A070103	Intermittent Capture		
					CT2407	Active implantable	E2403	No Clinical Signs, Symptoms or Conditions	A070103	Intermittent Capture		
					CT255	Transducers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070103	Intermittent Capture		

					CT2726	Physiological monitoring/measuring devices	E2403	No Clinical Signs, Symptoms or Conditions	A070103	Intermittent Capture		
					CT2727	Active implant devices/ancillaries	E2403	No Clinical Signs, Symptoms or Conditions	A070103	Intermittent Capture		
					CT2750	Implantable cardiac electrical stimulators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070103	Intermittent Capture		
					CT2885	Physiological monitoring/measuring system transmitters/receivers/transceivers	E2403	No Clinical Signs, Symptoms or Conditions	A070103	Intermittent Capture		
					CT2985	Active	E2403	No Clinical Signs, Symptoms or Conditions	A070103	Intermittent Capture		
					CT3001	Implantable metal	E2403	No Clinical Signs, Symptoms or Conditions	A070103	Intermittent Capture		
					CT3002	Implantable synthetic polymer	E2403	No Clinical Signs, Symptoms or Conditions	A070103	Intermittent Capture		
					CT3006	Implantable inorganic material	E2403	No Clinical Signs, Symptoms or Conditions	A070103	Intermittent Capture		
					CT314	Electrically powered	E2403	No Clinical Signs, Symptoms or Conditions	A070103	Intermittent Capture		
					CT321	Long-term surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A070103	Intermittent Capture		
					CT384	Sensors and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070103	Intermittent Capture		
					CT449	Generators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070103	Intermittent Capture		
					CT981	Single use	E2403	No Clinical Signs, Symptoms or Conditions	A070103	Intermittent Capture		
					CT983	Surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A070103	Intermittent Capture		
					CT994	Sensors	E2403	No Clinical Signs, Symptoms or Conditions	A070103	Intermittent Capture		
7431	2022	All other reportable incidents	47265	Dual-chamber implantable pacemaker, rate-responsive	CT141	Electrodes/Leads and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem	25/07/2022	25/07/2022
									A070504	Premature Discharge of Battery	25/07/2022	25/07/2022
							F08	Hospitalization or Prolonged Hospitalization	A0705	Battery Problem	25/07/2022	25/07/2022
									A070504	Premature Discharge of Battery	25/07/2022	25/07/2022
							F1905	Device Revision or Replacement	A0705	Battery Problem	25/07/2022	25/07/2022
									A070504	Premature Discharge of Battery	25/07/2022	25/07/2022
					CT145	Cardiovascular devices	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem	25/07/2022	25/07/2022
									A070504	Premature Discharge of Battery	25/07/2022	25/07/2022
							F08	Hospitalization or Prolonged Hospitalization	A0705	Battery Problem	25/07/2022	25/07/2022
									A070504	Premature Discharge of Battery	25/07/2022	25/07/2022
							F1905	Device Revision or Replacement	A0705	Battery Problem	25/07/2022	25/07/2022
									A070504	Premature Discharge of Battery	25/07/2022	25/07/2022
					CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem	25/07/2022	25/07/2022
									A070504	Premature Discharge of Battery	25/07/2022	25/07/2022
							F08	Hospitalization or Prolonged Hospitalization	A0705	Battery Problem	25/07/2022	25/07/2022
									A070504	Premature Discharge of Battery	25/07/2022	25/07/2022
							F1905	Device Revision or Replacement	A0705	Battery Problem	25/07/2022	25/07/2022
									A070504	Premature Discharge of Battery	25/07/2022	25/07/2022
					CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem	25/07/2022	25/07/2022
									A070504	Premature Discharge of Battery	25/07/2022	25/07/2022
							F08	Hospitalization or Prolonged Hospitalization	A0705	Battery Problem	25/07/2022	25/07/2022
									A070504	Premature Discharge of Battery	25/07/2022	25/07/2022
							F1905	Device Revision or Replacement	A0705	Battery Problem	25/07/2022	25/07/2022
									A070504	Premature Discharge of Battery	25/07/2022	25/07/2022
					CT233	Surgical	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem	25/07/2022	25/07/2022
									A070504	Premature Discharge of Battery	25/07/2022	25/07/2022
							F08	Hospitalization or Prolonged Hospitalization	A0705	Battery Problem	25/07/2022	25/07/2022
									A070504	Premature Discharge of Battery	25/07/2022	25/07/2022
							F1905	Device Revision or Replacement	A0705	Battery Problem	25/07/2022	25/07/2022
									A070504	Premature Discharge of Battery	25/07/2022	25/07/2022
					CT2406	Implantable	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem	25/07/2022	25/07/2022
									A070504	Premature Discharge of Battery	25/07/2022	25/07/2022
							F08	Hospitalization or Prolonged Hospitalization	A0705	Battery Problem	25/07/2022	25/07/2022
									A070504	Premature Discharge of Battery	25/07/2022	25/07/2022
							F1905	Device Revision or Replacement	A0705	Battery Problem	25/07/2022	25/07/2022
									A070504	Premature Discharge of Battery	25/07/2022	25/07/2022
					CT2407	Active implantable	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem	25/07/2022	25/07/2022
									A070504	Premature Discharge of Battery	25/07/2022	25/07/2022
							F08	Hospitalization or Prolonged Hospitalization	A0705	Battery Problem	25/07/2022	25/07/2022
									A070504	Premature Discharge of Battery	25/07/2022	25/07/2022
							F1905	Device Revision or Replacement	A0705	Battery Problem	25/07/2022	25/07/2022
									A070504	Premature Discharge of Battery	25/07/2022	25/07/2022
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem	25/07/2022	25/07/2022
									A070504	Premature Discharge of Battery	25/07/2022	25/07/2022
							F08	Hospitalization or Prolonged Hospitalization	A0705	Battery Problem	25/07/2022	25/07/2022
									A070504	Premature Discharge of Battery	25/07/2022	25/07/2022
							F1905	Device Revision or Replacement	A0705	Battery Problem	25/07/2022	25/07/2022
									A070504	Premature Discharge of Battery	25/07/2022	25/07/2022
					CT2727	Active implant devices/ancillaries	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem	25/07/2022	25/07/2022
									A070504	Premature Discharge of Battery	25/07/2022	25/07/2022
							F08	Hospitalization or Prolonged Hospitalization	A0705	Battery Problem	25/07/2022	25/07/2022
									A070504	Premature Discharge of Battery	25/07/2022	25/07/2022
							F1905	Device Revision or Replacement	A0705	Battery Problem	25/07/2022	25/07/2022
									A070504	Premature Discharge of Battery	25/07/2022	25/07/2022
					CT2750	Implantable cardiac electrical stimulators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem	25/07/2022	25/07/2022
									A070504	Premature Discharge of Battery	25/07/2022	25/07/2022
							F08	Hospitalization or Prolonged Hospitalization	A0705	Battery Problem	25/07/2022	25/07/2022
									A070504	Premature Discharge of Battery	25/07/2022	25/07/2022
							F1905	Device Revision or Replacement	A0705	Battery Problem	25/07/2022	25/07/2022
									A070504	Premature Discharge of Battery	25/07/2022	25/07/2022
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem	25/07/2022	25/07/2022
									A070504	Premature Discharge of Battery	25/07/2022	25/07/2022
							F08	Hospitalization or Prolonged Hospitalization	A0705	Battery Problem	25/07/2022	25/07/2022
									A070504	Premature Discharge of Battery	25/07/2022	25/07/2022
							F1905	Device Revision or Replacement	A0705	Battery Problem	25/07/2022	25/07/2022
									A070504	Premature Discharge of Battery	25/07/2022	25/07/2022
					CT2985	Active	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem	25/07/2022	25/07/2022
									A070504	Premature Discharge of Battery	25/07/2022	25/07/2022
							F08	Hospitalization or Prolonged Hospitalization	A0705	Battery Problem	25/07/2022	25/07/2022
									A070504	Premature Discharge of Battery	25/07/2022	25/07/2022
							F1905	Device Revision or Replacement	A0705	Battery Problem	25/07/2022	25/07/2022

									A070504	Premature Discharge of Battery	25/07/2022	25/07/2022
					CT3001	Implantable metal	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem	25/07/2022	25/07/2022
									A070504	Premature Discharge of Battery	25/07/2022	25/07/2022
							F08	Hospitalization or Prolonged Hospitalization	A0705	Battery Problem	25/07/2022	25/07/2022
									A070504	Premature Discharge of Battery	25/07/2022	25/07/2022
							F1905	Device Revision or Replacement	A0705	Battery Problem	25/07/2022	25/07/2022
									A070504	Premature Discharge of Battery	25/07/2022	25/07/2022
					CT3002	Implantable synthetic polymer	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem	25/07/2022	25/07/2022
									A070504	Premature Discharge of Battery	25/07/2022	25/07/2022
							F08	Hospitalization or Prolonged Hospitalization	A0705	Battery Problem	25/07/2022	25/07/2022
									A070504	Premature Discharge of Battery	25/07/2022	25/07/2022
							F1905	Device Revision or Replacement	A0705	Battery Problem	25/07/2022	25/07/2022
									A070504	Premature Discharge of Battery	25/07/2022	25/07/2022
					CT3006	Implantable inorganic material	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem	25/07/2022	25/07/2022
									A070504	Premature Discharge of Battery	25/07/2022	25/07/2022
							F08	Hospitalization or Prolonged Hospitalization	A0705	Battery Problem	25/07/2022	25/07/2022
									A070504	Premature Discharge of Battery	25/07/2022	25/07/2022
							F1905	Device Revision or Replacement	A0705	Battery Problem	25/07/2022	25/07/2022
									A070504	Premature Discharge of Battery	25/07/2022	25/07/2022
					CT314	Electrically powered	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem	25/07/2022	25/07/2022
									A070504	Premature Discharge of Battery	25/07/2022	25/07/2022
							F08	Hospitalization or Prolonged Hospitalization	A0705	Battery Problem	25/07/2022	25/07/2022
									A070504	Premature Discharge of Battery	25/07/2022	25/07/2022
							F1905	Device Revision or Replacement	A0705	Battery Problem	25/07/2022	25/07/2022
									A070504	Premature Discharge of Battery	25/07/2022	25/07/2022
					CT321	Long-term surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem	25/07/2022	25/07/2022
									A070504	Premature Discharge of Battery	25/07/2022	25/07/2022
							F08	Hospitalization or Prolonged Hospitalization	A0705	Battery Problem	25/07/2022	25/07/2022
									A070504	Premature Discharge of Battery	25/07/2022	25/07/2022
							F1905	Device Revision or Replacement	A0705	Battery Problem	25/07/2022	25/07/2022
									A070504	Premature Discharge of Battery	25/07/2022	25/07/2022
					CT449	Generators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem	25/07/2022	25/07/2022
									A070504	Premature Discharge of Battery	25/07/2022	25/07/2022
							F08	Hospitalization or Prolonged Hospitalization	A0705	Battery Problem	25/07/2022	25/07/2022
									A070504	Premature Discharge of Battery	25/07/2022	25/07/2022
							F1905	Device Revision or Replacement	A0705	Battery Problem	25/07/2022	25/07/2022
									A070504	Premature Discharge of Battery	25/07/2022	25/07/2022
					CT981	Single use	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem	25/07/2022	25/07/2022
									A070504	Premature Discharge of Battery	25/07/2022	25/07/2022
							F08	Hospitalization or Prolonged Hospitalization	A0705	Battery Problem	25/07/2022	25/07/2022
									A070504	Premature Discharge of Battery	25/07/2022	25/07/2022
							F1905	Device Revision or Replacement	A0705	Battery Problem	25/07/2022	25/07/2022
									A070504	Premature Discharge of Battery	25/07/2022	25/07/2022
					CT983	Surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem	25/07/2022	25/07/2022
									A070504	Premature Discharge of Battery	25/07/2022	25/07/2022
							F08	Hospitalization or Prolonged Hospitalization	A0705	Battery Problem	25/07/2022	25/07/2022
									A070504	Premature Discharge of Battery	25/07/2022	25/07/2022
							F1905	Device Revision or Replacement	A0705	Battery Problem	25/07/2022	25/07/2022
									A070504	Premature Discharge of Battery	25/07/2022	25/07/2022
7432	2021	All other reportable incidents	35852	Single-chamber implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	10/12/2021	10/12/2021
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	10/12/2021	10/12/2021
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	10/12/2021	10/12/2021
7433	2022	All other reportable incidents	47267	Single-chamber implantable pacemaker, rate-responsive	CT141	Electrodes/Leads and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	10/08/2022	10/08/2022
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	10/08/2022	10/08/2022
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	10/08/2022	10/08/2022
					CT145	Cardiovascular devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	10/08/2022	10/08/2022
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	10/08/2022	10/08/2022
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	10/08/2022	10/08/2022
					CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	10/08/2022	10/08/2022
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	10/08/2022	10/08/2022
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	10/08/2022	10/08/2022
					CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	10/08/2022	10/08/2022
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	10/08/2022	10/08/2022
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	10/08/2022	10/08/2022
					CT233	Surgical	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	10/08/2022	10/08/2022
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	10/08/2022	10/08/2022
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	10/08/2022	10/08/2022
					CT2406	Implantable	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	10/08/2022	10/08/2022
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	10/08/2022	10/08/2022
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	10/08/2022	10/08/2022
					CT2407	Active implantable	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	10/08/2022	10/08/2022
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	10/08/2022	10/08/2022
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	10/08/2022	10/08/2022
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	10/08/2022	10/08/2022
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	10/08/2022	10/08/2022
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	10/08/2022	10/08/2022
					CT2727	Active implant devices/ancillaries	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	10/08/2022	10/08/2022
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	10/08/2022	10/08/2022
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	10/08/2022	10/08/2022
					CT2750	Implantable cardiac electrical stimulators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	10/08/2022	10/08/2022
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	10/08/2022	10/08/2022
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	10/08/2022	10/08/2022
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	10/08/2022	10/08/2022
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	10/08/2022	10/08/2022
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	10/08/2022	10/08/2022
					CT2985	Active	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	10/08/2022	10/08/2022
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	10/08/2022	10/08/2022
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	10/08/2022	10/08/2022

					CT3001	Implantable metal	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	10/08/2022	10/08/2022
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	10/08/2022	10/08/2022
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	10/08/2022	10/08/2022
					CT3002	Implantable synthetic polymer	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	10/08/2022	10/08/2022
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	10/08/2022	10/08/2022
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	10/08/2022	10/08/2022
					CT3006	Implantable inorganic material	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	10/08/2022	10/08/2022
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	10/08/2022	10/08/2022
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	10/08/2022	10/08/2022
					CT314	Electrically powered	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	10/08/2022	10/08/2022
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	10/08/2022	10/08/2022
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	10/08/2022	10/08/2022
					CT321	Long-term surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	10/08/2022	10/08/2022
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	10/08/2022	10/08/2022
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	10/08/2022	10/08/2022
					CT449	Generators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	10/08/2022	10/08/2022
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	10/08/2022	10/08/2022
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	10/08/2022	10/08/2022
					CT981	Single use	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	10/08/2022	10/08/2022
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	10/08/2022	10/08/2022
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	10/08/2022	10/08/2022
					CT983	Surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	10/08/2022	10/08/2022
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	10/08/2022	10/08/2022
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	10/08/2022	10/08/2022
7434	2021	All other reportable incidents	17882	Physiologic-monitoring defibrillation system	CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070803	Failure to Power Up		
							F27	Problem Identified During Non-Clinical Procedure	A070803	Failure to Power Up		
7435	2021	All other reportable incidents	17882	Physiologic-monitoring defibrillation system	CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070803	Failure to Power Up		
							F27	Problem Identified During Non-Clinical Procedure	A070803	Failure to Power Up		
7436	2021	All other reportable incidents	17882	Physiologic-monitoring defibrillation system	CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A07	Electrical /Electronic Property Problem		
							F27	Problem Identified During Non-Clinical Procedure	A07	Electrical /Electronic Property Problem		
7437	2021	All other reportable incidents	17882	Physiologic-monitoring defibrillation system	CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A07	Electrical /Electronic Property Problem		
							F27	Problem Identified During Non-Clinical Procedure	A07	Electrical /Electronic Property Problem		
7438	2021	All other reportable incidents	17882	Physiologic-monitoring defibrillation system	CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A0709	Device Sensing Problem		
							F27	Problem Identified During Non-Clinical Procedure	A0709	Device Sensing Problem		
7439	2021	All other reportable incidents	35223	Endocardial/interventricular septal pacing lead	CT1684	Cardiac pacemakers and associated devices	E0112	Dizziness	A070909	Oversensing		
							E0601	Arrhythmia	A070909	Oversensing		
							E2403	No Clinical Signs, Symptoms or Conditions	A070909	Oversensing		
							F11	Non-Serious Injury/ Illness/ Impairment	A070909	Oversensing		
7440	2021	All other reportable incidents	35223	Endocardial/interventricular septal pacing lead	CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070901	Decreased Sensitivity		
									A070909	Oversensing		
									A072202	Low Impedance		
							F23	Unexpected Medical Intervention	A070901	Decreased Sensitivity		
									A070909	Oversensing		
									A072202	Low Impedance		
7441	2021	All other reportable incidents	17882	Physiologic-monitoring defibrillation system	CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A071301	Failure to Deliver Shock/Stimulation		
							F26	No Health Consequences or Impact	A071301	Failure to Deliver Shock/Stimulation		
7442	2021	All other reportable incidents	17882	Physiologic-monitoring defibrillation system	CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070803	Failure to Power Up		
							F27	Problem Identified During Non-Clinical Procedure	A070803	Failure to Power Up		
7443	2022	All other reportable incidents	35223	Endocardial/interventricular septal pacing lead	CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070909	Oversensing		
									A090801	Signal Artifact/Noise		
							F26	No Health Consequences or Impact	A070909	Oversensing		
									A090801	Signal Artifact/Noise		
7444	2022	All other reportable incidents	35223	Endocardial/interventricular septal pacing lead	CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A072202	Low Impedance		
							F26	No Health Consequences or Impact	A072202	Low Impedance		
7445	2021	All other reportable incidents	17882	Physiologic-monitoring defibrillation system	CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070912	Incorrect Interpretation of Signal		
							F26	No Health Consequences or Impact	A070912	Incorrect Interpretation of Signal		
7446	2021	All other reportable incidents	35852	Single-chamber implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A071201	Failure to Convert Rhythm		
							F08	Hospitalization or Prolonged Hospitalization	A071201	Failure to Convert Rhythm		
							F1905	Device Revision or Replacement	A071201	Failure to Convert Rhythm		
7447	2021	All other reportable incidents	47265	Dual-chamber implantable pacemaker, rate-responsive	CT141	Electrodes/Leads and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A0203	Defective Device	13/12/2021	13/12/2021
									A070504	Premature Discharge of Battery	13/12/2021	13/12/2021
									A0709	Device Sensing Problem	13/12/2021	13/12/2021
									A0722	Impedance Problem	13/12/2021	13/12/2021
							F08	Hospitalization or Prolonged Hospitalization	A0203	Defective Device	13/12/2021	13/12/2021
									A070504	Premature Discharge of Battery	13/12/2021	13/12/2021
									A0709	Device Sensing Problem	13/12/2021	13/12/2021
									A0722	Impedance Problem	13/12/2021	13/12/2021
							F1905	Device Revision or Replacement	A0203	Defective Device	13/12/2021	13/12/2021
									A070504	Premature Discharge of Battery	13/12/2021	13/12/2021
									A0709	Device Sensing Problem	13/12/2021	13/12/2021
									A0722	Impedance Problem	13/12/2021	13/12/2021
					CT145	Cardiovascular devices	E2403	No Clinical Signs, Symptoms or Conditions	A0203	Defective Device	13/12/2021	13/12/2021
									A070504	Premature Discharge of Battery	13/12/2021	13/12/2021
									A0709	Device Sensing Problem	13/12/2021	13/12/2021
									A0722	Impedance Problem	13/12/2021	13/12/2021
							F08	Hospitalization or Prolonged Hospitalization	A0203	Defective Device	13/12/2021	13/12/2021
									A070504	Premature Discharge of Battery	13/12/2021	13/12/2021
									A0709	Device Sensing Problem	13/12/2021	13/12/2021
									A0722	Impedance Problem	13/12/2021	13/12/2021
							F1905	Device Revision or Replacement	A0203	Defective Device	13/12/2021	13/12/2021
									A070504	Premature Discharge of Battery	13/12/2021	13/12/2021
									A0709	Device Sensing Problem	13/12/2021	13/12/2021
									A0722	Impedance Problem	13/12/2021	13/12/2021
					CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A0203	Defective Device	13/12/2021	13/12/2021
									A070504	Premature Discharge of Battery	13/12/2021	13/12/2021
									A0709	Device Sensing Problem	13/12/2021	13/12/2021
									A0722	Impedance Problem	13/12/2021	13/12/2021
							F08	Hospitalization or Prolonged Hospitalization	A0203	Defective Device	13/12/2021	13/12/2021

								A070504	Premature Discharge of Battery	13/12/2021	13/12/2021
								A0709	Device Sensing Problem	13/12/2021	13/12/2021
								A0722	Impedance Problem	13/12/2021	13/12/2021
						F1905	Device Revision or Replacement	A0203	Defective Device	13/12/2021	13/12/2021
								A070504	Premature Discharge of Battery	13/12/2021	13/12/2021
								A0709	Device Sensing Problem	13/12/2021	13/12/2021
								A0722	Impedance Problem	13/12/2021	13/12/2021
					CT1685	Cardiac pacemakers	E2403	A0203	Defective Device	13/12/2021	13/12/2021
								A070504	Premature Discharge of Battery	13/12/2021	13/12/2021
								A0709	Device Sensing Problem	13/12/2021	13/12/2021
								A0722	Impedance Problem	13/12/2021	13/12/2021
							F08	A0203	Defective Device	13/12/2021	13/12/2021
								A070504	Premature Discharge of Battery	13/12/2021	13/12/2021
								A0709	Device Sensing Problem	13/12/2021	13/12/2021
								A0722	Impedance Problem	13/12/2021	13/12/2021
							F1905	A0203	Defective Device	13/12/2021	13/12/2021
								A070504	Premature Discharge of Battery	13/12/2021	13/12/2021
								A0709	Device Sensing Problem	13/12/2021	13/12/2021
								A0722	Impedance Problem	13/12/2021	13/12/2021
					CT233	Surgical	E2403	A0203	Defective Device	13/12/2021	13/12/2021
								A070504	Premature Discharge of Battery	13/12/2021	13/12/2021
								A0709	Device Sensing Problem	13/12/2021	13/12/2021
								A0722	Impedance Problem	13/12/2021	13/12/2021
							F08	A0203	Defective Device	13/12/2021	13/12/2021
								A070504	Premature Discharge of Battery	13/12/2021	13/12/2021
								A0709	Device Sensing Problem	13/12/2021	13/12/2021
								A0722	Impedance Problem	13/12/2021	13/12/2021
							F1905	A0203	Defective Device	13/12/2021	13/12/2021
								A070504	Premature Discharge of Battery	13/12/2021	13/12/2021
								A0709	Device Sensing Problem	13/12/2021	13/12/2021
								A0722	Impedance Problem	13/12/2021	13/12/2021
					CT2406	Implantable	E2403	A0203	Defective Device	13/12/2021	13/12/2021
								A070504	Premature Discharge of Battery	13/12/2021	13/12/2021
								A0709	Device Sensing Problem	13/12/2021	13/12/2021
								A0722	Impedance Problem	13/12/2021	13/12/2021
							F08	A0203	Defective Device	13/12/2021	13/12/2021
								A070504	Premature Discharge of Battery	13/12/2021	13/12/2021
								A0709	Device Sensing Problem	13/12/2021	13/12/2021
								A0722	Impedance Problem	13/12/2021	13/12/2021
							F1905	A0203	Defective Device	13/12/2021	13/12/2021
								A070504	Premature Discharge of Battery	13/12/2021	13/12/2021
								A0709	Device Sensing Problem	13/12/2021	13/12/2021
								A0722	Impedance Problem	13/12/2021	13/12/2021
					CT2407	Active implantable	E2403	A0203	Defective Device	13/12/2021	13/12/2021
								A070504	Premature Discharge of Battery	13/12/2021	13/12/2021
								A0709	Device Sensing Problem	13/12/2021	13/12/2021
								A0722	Impedance Problem	13/12/2021	13/12/2021
							F08	A0203	Defective Device	13/12/2021	13/12/2021
								A070504	Premature Discharge of Battery	13/12/2021	13/12/2021
								A0709	Device Sensing Problem	13/12/2021	13/12/2021
								A0722	Impedance Problem	13/12/2021	13/12/2021
							F1905	A0203	Defective Device	13/12/2021	13/12/2021
								A070504	Premature Discharge of Battery	13/12/2021	13/12/2021
								A0709	Device Sensing Problem	13/12/2021	13/12/2021
								A0722	Impedance Problem	13/12/2021	13/12/2021
					CT257	Implantable cardiac pacemakers	E2403	A0203	Defective Device	13/12/2021	13/12/2021
								A070504	Premature Discharge of Battery	13/12/2021	13/12/2021
								A0709	Device Sensing Problem	13/12/2021	13/12/2021
								A0722	Impedance Problem	13/12/2021	13/12/2021
							F08	A0203	Defective Device	13/12/2021	13/12/2021
								A070504	Premature Discharge of Battery	13/12/2021	13/12/2021
								A0709	Device Sensing Problem	13/12/2021	13/12/2021
								A0722	Impedance Problem	13/12/2021	13/12/2021
							F1905	A0203	Defective Device	13/12/2021	13/12/2021
								A070504	Premature Discharge of Battery	13/12/2021	13/12/2021
								A0709	Device Sensing Problem	13/12/2021	13/12/2021
								A0722	Impedance Problem	13/12/2021	13/12/2021
					CT2727	Active implant devices/ancillaries	E2403	A0203	Defective Device	13/12/2021	13/12/2021
								A070504	Premature Discharge of Battery	13/12/2021	13/12/2021
								A0709	Device Sensing Problem	13/12/2021	13/12/2021
								A0722	Impedance Problem	13/12/2021	13/12/2021
							F08	A0203	Defective Device	13/12/2021	13/12/2021
								A070504	Premature Discharge of Battery	13/12/2021	13/12/2021
								A0709	Device Sensing Problem	13/12/2021	13/12/2021
								A0722	Impedance Problem	13/12/2021	13/12/2021
							F1905	A0203	Defective Device	13/12/2021	13/12/2021
								A070504	Premature Discharge of Battery	13/12/2021	13/12/2021
								A0709	Device Sensing Problem	13/12/2021	13/12/2021
								A0722	Impedance Problem	13/12/2021	13/12/2021
					CT2750	Implantable cardiac electrical stimulators and associated devices	E2403	A0203	Defective Device	13/12/2021	13/12/2021
								A070504	Premature Discharge of Battery	13/12/2021	13/12/2021
								A0709	Device Sensing Problem	13/12/2021	13/12/2021
								A0722	Impedance Problem	13/12/2021	13/12/2021
							F08	A0203	Defective Device	13/12/2021	13/12/2021
								A070504	Premature Discharge of Battery	13/12/2021	13/12/2021
								A0709	Device Sensing Problem	13/12/2021	13/12/2021
								A0722	Impedance Problem	13/12/2021	13/12/2021
							F1905	A0203	Defective Device	13/12/2021	13/12/2021

								A070504	Premature Discharge of Battery	13/12/2021	13/12/2021
								A0709	Device Sensing Problem	13/12/2021	13/12/2021
								A0722	Impedance Problem	13/12/2021	13/12/2021
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A0203	Defective Device	13/12/2021	13/12/2021
									A070504	Premature Discharge of Battery	13/12/2021	13/12/2021
									A0709	Device Sensing Problem	13/12/2021	13/12/2021
									A0722	Impedance Problem	13/12/2021	13/12/2021
							F08	Hospitalization or Prolonged Hospitalization	A0203	Defective Device	13/12/2021	13/12/2021
									A070504	Premature Discharge of Battery	13/12/2021	13/12/2021
									A0709	Device Sensing Problem	13/12/2021	13/12/2021
									A0722	Impedance Problem	13/12/2021	13/12/2021
							F1905	Device Revision or Replacement	A0203	Defective Device	13/12/2021	13/12/2021
									A070504	Premature Discharge of Battery	13/12/2021	13/12/2021
									A0709	Device Sensing Problem	13/12/2021	13/12/2021
									A0722	Impedance Problem	13/12/2021	13/12/2021
					CT2985	Active	E2403	No Clinical Signs, Symptoms or Conditions	A0203	Defective Device	13/12/2021	13/12/2021
									A070504	Premature Discharge of Battery	13/12/2021	13/12/2021
									A0709	Device Sensing Problem	13/12/2021	13/12/2021
									A0722	Impedance Problem	13/12/2021	13/12/2021
							F08	Hospitalization or Prolonged Hospitalization	A0203	Defective Device	13/12/2021	13/12/2021
									A070504	Premature Discharge of Battery	13/12/2021	13/12/2021
									A0709	Device Sensing Problem	13/12/2021	13/12/2021
									A0722	Impedance Problem	13/12/2021	13/12/2021
							F1905	Device Revision or Replacement	A0203	Defective Device	13/12/2021	13/12/2021
									A070504	Premature Discharge of Battery	13/12/2021	13/12/2021
									A0709	Device Sensing Problem	13/12/2021	13/12/2021
									A0722	Impedance Problem	13/12/2021	13/12/2021
					CT3001	Implantable metal	E2403	No Clinical Signs, Symptoms or Conditions	A0203	Defective Device	13/12/2021	13/12/2021
									A070504	Premature Discharge of Battery	13/12/2021	13/12/2021
									A0709	Device Sensing Problem	13/12/2021	13/12/2021
									A0722	Impedance Problem	13/12/2021	13/12/2021
							F08	Hospitalization or Prolonged Hospitalization	A0203	Defective Device	13/12/2021	13/12/2021
									A070504	Premature Discharge of Battery	13/12/2021	13/12/2021
									A0709	Device Sensing Problem	13/12/2021	13/12/2021
									A0722	Impedance Problem	13/12/2021	13/12/2021
							F1905	Device Revision or Replacement	A0203	Defective Device	13/12/2021	13/12/2021
									A070504	Premature Discharge of Battery	13/12/2021	13/12/2021
									A0709	Device Sensing Problem	13/12/2021	13/12/2021
									A0722	Impedance Problem	13/12/2021	13/12/2021
					CT3002	Implantable synthetic polymer	E2403	No Clinical Signs, Symptoms or Conditions	A0203	Defective Device	13/12/2021	13/12/2021
									A070504	Premature Discharge of Battery	13/12/2021	13/12/2021
									A0709	Device Sensing Problem	13/12/2021	13/12/2021
									A0722	Impedance Problem	13/12/2021	13/12/2021
							F08	Hospitalization or Prolonged Hospitalization	A0203	Defective Device	13/12/2021	13/12/2021
									A070504	Premature Discharge of Battery	13/12/2021	13/12/2021
									A0709	Device Sensing Problem	13/12/2021	13/12/2021
									A0722	Impedance Problem	13/12/2021	13/12/2021
							F1905	Device Revision or Replacement	A0203	Defective Device	13/12/2021	13/12/2021
									A070504	Premature Discharge of Battery	13/12/2021	13/12/2021
									A0709	Device Sensing Problem	13/12/2021	13/12/2021
									A0722	Impedance Problem	13/12/2021	13/12/2021
					CT3006	Implantable inorganic material	E2403	No Clinical Signs, Symptoms or Conditions	A0203	Defective Device	13/12/2021	13/12/2021
									A070504	Premature Discharge of Battery	13/12/2021	13/12/2021
									A0709	Device Sensing Problem	13/12/2021	13/12/2021
									A0722	Impedance Problem	13/12/2021	13/12/2021
							F08	Hospitalization or Prolonged Hospitalization	A0203	Defective Device	13/12/2021	13/12/2021
									A070504	Premature Discharge of Battery	13/12/2021	13/12/2021
									A0709	Device Sensing Problem	13/12/2021	13/12/2021
									A0722	Impedance Problem	13/12/2021	13/12/2021
							F1905	Device Revision or Replacement	A0203	Defective Device	13/12/2021	13/12/2021
									A070504	Premature Discharge of Battery	13/12/2021	13/12/2021
									A0709	Device Sensing Problem	13/12/2021	13/12/2021
									A0722	Impedance Problem	13/12/2021	13/12/2021
					CT314	Electrically powered	E2403	No Clinical Signs, Symptoms or Conditions	A0203	Defective Device	13/12/2021	13/12/2021
									A070504	Premature Discharge of Battery	13/12/2021	13/12/2021
									A0709	Device Sensing Problem	13/12/2021	13/12/2021
									A0722	Impedance Problem	13/12/2021	13/12/2021
							F08	Hospitalization or Prolonged Hospitalization	A0203	Defective Device	13/12/2021	13/12/2021
									A070504	Premature Discharge of Battery	13/12/2021	13/12/2021
									A0709	Device Sensing Problem	13/12/2021	13/12/2021
									A0722	Impedance Problem	13/12/2021	13/12/2021
							F1905	Device Revision or Replacement	A0203	Defective Device	13/12/2021	13/12/2021
									A070504	Premature Discharge of Battery	13/12/2021	13/12/2021
									A0709	Device Sensing Problem	13/12/2021	13/12/2021
									A0722	Impedance Problem	13/12/2021	13/12/2021
					CT321	Long-term surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A0203	Defective Device	13/12/2021	13/12/2021
									A070504	Premature Discharge of Battery	13/12/2021	13/12/2021
									A0709	Device Sensing Problem	13/12/2021	13/12/2021
									A0722	Impedance Problem	13/12/2021	13/12/2021
							F08	Hospitalization or Prolonged Hospitalization	A0203	Defective Device	13/12/2021	13/12/2021
									A070504	Premature Discharge of Battery	13/12/2021	13/12/2021
									A0709	Device Sensing Problem	13/12/2021	13/12/2021
									A0722	Impedance Problem	13/12/2021	13/12/2021
							F1905	Device Revision or Replacement	A0203	Defective Device	13/12/2021	13/12/2021
									A070504	Premature Discharge of Battery	13/12/2021	13/12/2021
									A0709	Device Sensing Problem	13/12/2021	13/12/2021
									A0722	Impedance Problem	13/12/2021	13/12/2021
					CT449	Generators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A0203	Defective Device	13/12/2021	13/12/2021

								A070504	Premature Discharge of Battery	13/12/2021	13/12/2021
								A0709	Device Sensing Problem	13/12/2021	13/12/2021
								A0722	Impedance Problem	13/12/2021	13/12/2021
						F08	Hospitalization or Prolonged Hospitalization	A0203	Defective Device	13/12/2021	13/12/2021
								A070504	Premature Discharge of Battery	13/12/2021	13/12/2021
								A0709	Device Sensing Problem	13/12/2021	13/12/2021
								A0722	Impedance Problem	13/12/2021	13/12/2021
						F1905	Device Revision or Replacement	A0203	Defective Device	13/12/2021	13/12/2021
								A070504	Premature Discharge of Battery	13/12/2021	13/12/2021
								A0709	Device Sensing Problem	13/12/2021	13/12/2021
								A0722	Impedance Problem	13/12/2021	13/12/2021
				CT981	Single use	E2403	No Clinical Signs, Symptoms or Conditions	A0203	Defective Device	13/12/2021	13/12/2021
								A070504	Premature Discharge of Battery	13/12/2021	13/12/2021
								A0709	Device Sensing Problem	13/12/2021	13/12/2021
								A0722	Impedance Problem	13/12/2021	13/12/2021
						F08	Hospitalization or Prolonged Hospitalization	A0203	Defective Device	13/12/2021	13/12/2021
								A070504	Premature Discharge of Battery	13/12/2021	13/12/2021
								A0709	Device Sensing Problem	13/12/2021	13/12/2021
								A0722	Impedance Problem	13/12/2021	13/12/2021
						F1905	Device Revision or Replacement	A0203	Defective Device	13/12/2021	13/12/2021
								A070504	Premature Discharge of Battery	13/12/2021	13/12/2021
								A0709	Device Sensing Problem	13/12/2021	13/12/2021
								A0722	Impedance Problem	13/12/2021	13/12/2021
				CT983	Surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A0203	Defective Device	13/12/2021	13/12/2021
								A070504	Premature Discharge of Battery	13/12/2021	13/12/2021
								A0709	Device Sensing Problem	13/12/2021	13/12/2021
								A0722	Impedance Problem	13/12/2021	13/12/2021
						F08	Hospitalization or Prolonged Hospitalization	A0203	Defective Device	13/12/2021	13/12/2021
								A070504	Premature Discharge of Battery	13/12/2021	13/12/2021
								A0709	Device Sensing Problem	13/12/2021	13/12/2021
								A0722	Impedance Problem	13/12/2021	13/12/2021
						F1905	Device Revision or Replacement	A0203	Defective Device	13/12/2021	13/12/2021
								A070504	Premature Discharge of Battery	13/12/2021	13/12/2021
								A0709	Device Sensing Problem	13/12/2021	13/12/2021
								A0722	Impedance Problem	13/12/2021	13/12/2021
7448	2022	All other reportable incidents	47267	Single-chamber implantable pacemaker, rate-responsive	CT141	Electrodes/Leads and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT145	Cardiovascular devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT233	Surgical	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT2406	Implantable	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT2407	Active implantable	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT2727	Active implant devices/ancillaries	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT2750	Implantable cardiac electrical stimulators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT2985	Active	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT3001	Implantable metal	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT3002	Implantable synthetic polymer	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT3006	Implantable inorganic material	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT314	Electrically powered	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT321	Long-term surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT449	Generators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT981	Single use	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT983	Surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
7449	2022	All other reportable incidents	36241	Pacemaker/defibrillator lead	CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A05	Mechanical Problem	30/03/2022	30/03/2022
							F1908	Prolonged Surgery	A05	Mechanical Problem	30/03/2022	30/03/2022
7450	2022	All other reportable incidents	47263	Cardiac resynchronization therapy implantable pacemaker	CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A010402	Migration		
									A070909	Oversensing		
									A072201	High Impedance		
							F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A010402	Migration		
									A070909	Oversensing		
									A072201	High Impedance		
							F2203	Imaging Required	A010402	Migration		
									A070909	Oversensing		
									A072201	High Impedance		
					CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A010402	Migration		
									A070909	Oversensing		

									A072201	High Impedance		
							F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A010402	Migration		
									A070909	Oversensing		
									A072201	High Impedance		
							F2203	Imaging Required	A010402	Migration		
									A070909	Oversensing		
									A072201	High Impedance		
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A010402	Migration		
									A070909	Oversensing		
									A072201	High Impedance		
							F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A010402	Migration		
									A070909	Oversensing		
									A072201	High Impedance		
							F2203	Imaging Required	A010402	Migration		
									A070909	Oversensing		
									A072201	High Impedance		
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A010402	Migration		
									A070909	Oversensing		
									A072201	High Impedance		
							F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A010402	Migration		
									A070909	Oversensing		
									A072201	High Impedance		
							F2203	Imaging Required	A010402	Migration		
									A070909	Oversensing		
									A072201	High Impedance		
7451	2022	All other reportable incidents	36241	Pacemaker/defibrillator lead	CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A0203	Defective Device		
									A072201	High Impedance		
							F2203	Imaging Required	A0203	Defective Device		
									A072201	High Impedance		
7452	2021	All other reportable incidents	47265	Dual-chamber implantable pacemaker, rate-responsive	CT141	Electrodes/Leads and associated devices	E0112	Dizziness	A070504	Premature Discharge of Battery	21/12/2021	21/12/2021
									A070909	Oversensing	21/12/2021	21/12/2021
							E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	21/12/2021	21/12/2021
									A070909	Oversensing	21/12/2021	21/12/2021
							F1901	Additional Surgery	A070504	Premature Discharge of Battery	21/12/2021	21/12/2021
									A070909	Oversensing	21/12/2021	21/12/2021
					CT145	Cardiovascular devices	E0112	Dizziness	A070504	Premature Discharge of Battery	21/12/2021	21/12/2021
									A070909	Oversensing	21/12/2021	21/12/2021
							E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	21/12/2021	21/12/2021
									A070909	Oversensing	21/12/2021	21/12/2021
							F1901	Additional Surgery	A070504	Premature Discharge of Battery	21/12/2021	21/12/2021
									A070909	Oversensing	21/12/2021	21/12/2021
					CT1684	Cardiac pacemakers and associated devices	E0112	Dizziness	A070504	Premature Discharge of Battery	21/12/2021	21/12/2021
									A070909	Oversensing	21/12/2021	21/12/2021
							E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	21/12/2021	21/12/2021
									A070909	Oversensing	21/12/2021	21/12/2021
							F1901	Additional Surgery	A070504	Premature Discharge of Battery	21/12/2021	21/12/2021
									A070909	Oversensing	21/12/2021	21/12/2021
					CT1685	Cardiac pacemakers	E0112	Dizziness	A070504	Premature Discharge of Battery	21/12/2021	21/12/2021
									A070909	Oversensing	21/12/2021	21/12/2021
							E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	21/12/2021	21/12/2021
									A070909	Oversensing	21/12/2021	21/12/2021
							F1901	Additional Surgery	A070504	Premature Discharge of Battery	21/12/2021	21/12/2021
									A070909	Oversensing	21/12/2021	21/12/2021
					CT233	Surgical	E0112	Dizziness	A070504	Premature Discharge of Battery	21/12/2021	21/12/2021
									A070909	Oversensing	21/12/2021	21/12/2021
							E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	21/12/2021	21/12/2021
									A070909	Oversensing	21/12/2021	21/12/2021
							F1901	Additional Surgery	A070504	Premature Discharge of Battery	21/12/2021	21/12/2021
									A070909	Oversensing	21/12/2021	21/12/2021
					CT2406	Implantable	E0112	Dizziness	A070504	Premature Discharge of Battery	21/12/2021	21/12/2021
									A070909	Oversensing	21/12/2021	21/12/2021
							E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	21/12/2021	21/12/2021
									A070909	Oversensing	21/12/2021	21/12/2021
							F1901	Additional Surgery	A070504	Premature Discharge of Battery	21/12/2021	21/12/2021
									A070909	Oversensing	21/12/2021	21/12/2021
					CT2407	Active implantable	E0112	Dizziness	A070504	Premature Discharge of Battery	21/12/2021	21/12/2021
									A070909	Oversensing	21/12/2021	21/12/2021
							E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	21/12/2021	21/12/2021
									A070909	Oversensing	21/12/2021	21/12/2021
							F1901	Additional Surgery	A070504	Premature Discharge of Battery	21/12/2021	21/12/2021
									A070909	Oversensing	21/12/2021	21/12/2021
					CT257	Implantable cardiac pacemakers	E0112	Dizziness	A070504	Premature Discharge of Battery	21/12/2021	21/12/2021
									A070909	Oversensing	21/12/2021	21/12/2021
							E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	21/12/2021	21/12/2021
									A070909	Oversensing	21/12/2021	21/12/2021
							F1901	Additional Surgery	A070504	Premature Discharge of Battery	21/12/2021	21/12/2021
									A070909	Oversensing	21/12/2021	21/12/2021
					CT2727	Active implant devices/ancillaries	E0112	Dizziness	A070504	Premature Discharge of Battery	21/12/2021	21/12/2021
									A070909	Oversensing	21/12/2021	21/12/2021
							E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	21/12/2021	21/12/2021
									A070909	Oversensing	21/12/2021	21/12/2021
							F1901	Additional Surgery	A070504	Premature Discharge of Battery	21/12/2021	21/12/2021
									A070909	Oversensing	21/12/2021	21/12/2021
					CT2750	Implantable cardiac electrical stimulators and associated devices	E0112	Dizziness	A070504	Premature Discharge of Battery	21/12/2021	21/12/2021
									A070909	Oversensing	21/12/2021	21/12/2021
							E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	21/12/2021	21/12/2021
									A070909	Oversensing	21/12/2021	21/12/2021
							F1901	Additional Surgery	A070504	Premature Discharge of Battery	21/12/2021	21/12/2021

									A070909	Oversensing	21/12/2021	21/12/2021
					CT2866	Implantable cardiac electrical stimulators	E0112	Dizziness	A070504	Premature Discharge of Battery	21/12/2021	21/12/2021
									A070909	Oversensing	21/12/2021	21/12/2021
							E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	21/12/2021	21/12/2021
									A070909	Oversensing	21/12/2021	21/12/2021
							F1901	Additional Surgery	A070504	Premature Discharge of Battery	21/12/2021	21/12/2021
									A070909	Oversensing	21/12/2021	21/12/2021
					CT2985	Active	E0112	Dizziness	A070504	Premature Discharge of Battery	21/12/2021	21/12/2021
									A070909	Oversensing	21/12/2021	21/12/2021
							E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	21/12/2021	21/12/2021
									A070909	Oversensing	21/12/2021	21/12/2021
							F1901	Additional Surgery	A070504	Premature Discharge of Battery	21/12/2021	21/12/2021
									A070909	Oversensing	21/12/2021	21/12/2021
					CT3001	Implantable metal	E0112	Dizziness	A070504	Premature Discharge of Battery	21/12/2021	21/12/2021
									A070909	Oversensing	21/12/2021	21/12/2021
							E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	21/12/2021	21/12/2021
									A070909	Oversensing	21/12/2021	21/12/2021
							F1901	Additional Surgery	A070504	Premature Discharge of Battery	21/12/2021	21/12/2021
									A070909	Oversensing	21/12/2021	21/12/2021
					CT3002	Implantable synthetic polymer	E0112	Dizziness	A070504	Premature Discharge of Battery	21/12/2021	21/12/2021
									A070909	Oversensing	21/12/2021	21/12/2021
							E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	21/12/2021	21/12/2021
									A070909	Oversensing	21/12/2021	21/12/2021
							F1901	Additional Surgery	A070504	Premature Discharge of Battery	21/12/2021	21/12/2021
									A070909	Oversensing	21/12/2021	21/12/2021
					CT3006	Implantable inorganic material	E0112	Dizziness	A070504	Premature Discharge of Battery	21/12/2021	21/12/2021
									A070909	Oversensing	21/12/2021	21/12/2021
							E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	21/12/2021	21/12/2021
									A070909	Oversensing	21/12/2021	21/12/2021
							F1901	Additional Surgery	A070504	Premature Discharge of Battery	21/12/2021	21/12/2021
									A070909	Oversensing	21/12/2021	21/12/2021
					CT314	Electrically powered	E0112	Dizziness	A070504	Premature Discharge of Battery	21/12/2021	21/12/2021
									A070909	Oversensing	21/12/2021	21/12/2021
							E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	21/12/2021	21/12/2021
									A070909	Oversensing	21/12/2021	21/12/2021
							F1901	Additional Surgery	A070504	Premature Discharge of Battery	21/12/2021	21/12/2021
									A070909	Oversensing	21/12/2021	21/12/2021
					CT321	Long-term surgical invasive	E0112	Dizziness	A070504	Premature Discharge of Battery	21/12/2021	21/12/2021
									A070909	Oversensing	21/12/2021	21/12/2021
							E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	21/12/2021	21/12/2021
									A070909	Oversensing	21/12/2021	21/12/2021
							F1901	Additional Surgery	A070504	Premature Discharge of Battery	21/12/2021	21/12/2021
									A070909	Oversensing	21/12/2021	21/12/2021
					CT449	Generators and associated devices	E0112	Dizziness	A070504	Premature Discharge of Battery	21/12/2021	21/12/2021
									A070909	Oversensing	21/12/2021	21/12/2021
							E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	21/12/2021	21/12/2021
									A070909	Oversensing	21/12/2021	21/12/2021
							F1901	Additional Surgery	A070504	Premature Discharge of Battery	21/12/2021	21/12/2021
									A070909	Oversensing	21/12/2021	21/12/2021
					CT981	Single use	E0112	Dizziness	A070504	Premature Discharge of Battery	21/12/2021	21/12/2021
									A070909	Oversensing	21/12/2021	21/12/2021
							E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	21/12/2021	21/12/2021
									A070909	Oversensing	21/12/2021	21/12/2021
							F1901	Additional Surgery	A070504	Premature Discharge of Battery	21/12/2021	21/12/2021
									A070909	Oversensing	21/12/2021	21/12/2021
					CT983	Surgical invasive	E0112	Dizziness	A070504	Premature Discharge of Battery	21/12/2021	21/12/2021
									A070909	Oversensing	21/12/2021	21/12/2021
							E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	21/12/2021	21/12/2021
									A070909	Oversensing	21/12/2021	21/12/2021
							F1901	Additional Surgery	A070504	Premature Discharge of Battery	21/12/2021	21/12/2021
									A070909	Oversensing	21/12/2021	21/12/2021
7453	2021	All other reportable incidents	36241	Pacemaker/defibrillator lead	CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A051201	Device Dislodged or Dislocated		
							F26	No Health Consequences or Impact	A051201	Device Dislodged or Dislocated		
7454	2021	All other reportable incidents	37265	Dual-chamber implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E0601	Arrhythmia	A070910	Undersensing		
									A0712	Pacing Problem		
									A071202	Inaccurate Synchronization		
							E2403	No Clinical Signs, Symptoms or Conditions	A070910	Undersensing		
									A0712	Pacing Problem		
									A071202	Inaccurate Synchronization		
							F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A070910	Undersensing		
									A0712	Pacing Problem		
									A071202	Inaccurate Synchronization		
							F15	Recognized Device or Procedural Complication	A070910	Undersensing		
									A0712	Pacing Problem		
									A071202	Inaccurate Synchronization		
7455	2021	All other reportable incidents	36241	Pacemaker/defibrillator lead	CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070102	High Capture Threshold	10/11/2021	10/11/2021
							F08	Hospitalization or Prolonged Hospitalization	A070102	High Capture Threshold	10/11/2021	10/11/2021
							F1905	Device Revision or Replacement	A070102	High Capture Threshold	10/11/2021	10/11/2021
7456	2021	All other reportable incidents	47265	Dual-chamber implantable pacemaker, rate-responsive	CT141	Electrodes/Leads and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT145	Cardiovascular devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT233	Surgical	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		

					CT2406	Implantable	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT2407	Active implantable	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT2727	Active implant devices/ancillaries	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT2750	Implantable cardiac electrical stimulators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT2985	Active	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT3001	Implantable metal	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT3002	Implantable synthetic polymer	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT3006	Implantable inorganic material	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT314	Electrically powered	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT321	Long-term surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT449	Generators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT981	Single use	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT983	Surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
7457	2022	All other reportable incidents	47269	Transtelephonic implantable pacemaker analysis system	CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A090805	Incorrect Measurement		
							F26	No Health Consequences or Impact	A090805	Incorrect Measurement		
7458	2021	All other reportable incidents	47265	Dual-chamber implantable pacemaker, rate-responsive	CT141	Electrodes/Leads and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
					CT145	Cardiovascular devices	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
					CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
					CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
					CT233	Surgical	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
					CT2406	Implantable	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
					CT2407	Active implantable	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
					CT2727	Active implant devices/ancillaries	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
					CT2750	Implantable cardiac electrical stimulators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
					CT2985	Active	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
					CT3001	Implantable metal	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
					CT3002	Implantable synthetic polymer	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		

					CT3006	Implantable inorganic material	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
					CT314	Electrically powered	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
					CT321	Long-term surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
					CT449	Generators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
					CT981	Single use	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
					CT983	Surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
7459	2021	All other reportable incidents	47265	Dual-chamber implantable pacemaker, rate-responsive	CT141	Electrodes/Leads and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	07/12/2021	07/12/2021
							F1901	Additional Surgery	A070504	Premature Discharge of Battery	07/12/2021	07/12/2021
					CT145	Cardiovascular devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	07/12/2021	07/12/2021
							F1901	Additional Surgery	A070504	Premature Discharge of Battery	07/12/2021	07/12/2021
					CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	07/12/2021	07/12/2021
							F1901	Additional Surgery	A070504	Premature Discharge of Battery	07/12/2021	07/12/2021
					CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	07/12/2021	07/12/2021
							F1901	Additional Surgery	A070504	Premature Discharge of Battery	07/12/2021	07/12/2021
					CT233	Surgical	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	07/12/2021	07/12/2021
							F1901	Additional Surgery	A070504	Premature Discharge of Battery	07/12/2021	07/12/2021
					CT2406	Implantable	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	07/12/2021	07/12/2021
							F1901	Additional Surgery	A070504	Premature Discharge of Battery	07/12/2021	07/12/2021
					CT2407	Active implantable	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	07/12/2021	07/12/2021
							F1901	Additional Surgery	A070504	Premature Discharge of Battery	07/12/2021	07/12/2021
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	07/12/2021	07/12/2021
							F1901	Additional Surgery	A070504	Premature Discharge of Battery	07/12/2021	07/12/2021
					CT2727	Active implant devices/ancillaries	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	07/12/2021	07/12/2021
							F1901	Additional Surgery	A070504	Premature Discharge of Battery	07/12/2021	07/12/2021
					CT2750	Implantable cardiac electrical stimulators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	07/12/2021	07/12/2021
							F1901	Additional Surgery	A070504	Premature Discharge of Battery	07/12/2021	07/12/2021
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	07/12/2021	07/12/2021
							F1901	Additional Surgery	A070504	Premature Discharge of Battery	07/12/2021	07/12/2021
					CT2985	Active	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	07/12/2021	07/12/2021
							F1901	Additional Surgery	A070504	Premature Discharge of Battery	07/12/2021	07/12/2021
					CT3001	Implantable metal	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	07/12/2021	07/12/2021
							F1901	Additional Surgery	A070504	Premature Discharge of Battery	07/12/2021	07/12/2021
					CT3002	Implantable synthetic polymer	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	07/12/2021	07/12/2021
							F1901	Additional Surgery	A070504	Premature Discharge of Battery	07/12/2021	07/12/2021
					CT3006	Implantable inorganic material	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	07/12/2021	07/12/2021
							F1901	Additional Surgery	A070504	Premature Discharge of Battery	07/12/2021	07/12/2021
					CT314	Electrically powered	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	07/12/2021	07/12/2021
							F1901	Additional Surgery	A070504	Premature Discharge of Battery	07/12/2021	07/12/2021
					CT321	Long-term surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	07/12/2021	07/12/2021
							F1901	Additional Surgery	A070504	Premature Discharge of Battery	07/12/2021	07/12/2021
					CT449	Generators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	07/12/2021	07/12/2021
							F1901	Additional Surgery	A070504	Premature Discharge of Battery	07/12/2021	07/12/2021
					CT981	Single use	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	07/12/2021	07/12/2021
							F1901	Additional Surgery	A070504	Premature Discharge of Battery	07/12/2021	07/12/2021
					CT983	Surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	07/12/2021	07/12/2021
							F1901	Additional Surgery	A070504	Premature Discharge of Battery	07/12/2021	07/12/2021
7460	2022	All other reportable incidents	37265	Dual-chamber implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
7461	2021	All other reportable incidents	35852	Single-chamber implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem	22/12/2021	22/12/2021
									A070504	Premature Discharge of Battery	22/12/2021	22/12/2021
							F08	Hospitalization or Prolonged Hospitalization	A0705	Battery Problem	22/12/2021	22/12/2021
									A070504	Premature Discharge of Battery	22/12/2021	22/12/2021
							F1905	Device Revision or Replacement	A0705	Battery Problem	22/12/2021	22/12/2021
									A070504	Premature Discharge of Battery	22/12/2021	22/12/2021
7462	2022	All other reportable incidents	63188	Left-ventricular wireless pacing system sensor/transmitter	CT141	Electrodes/Leads and associated devices	E1716	Skin Discoloration	A010402	Migration	25/07/2022	25/07/2022
							E2403	No Clinical Signs, Symptoms or Conditions	A010402	Migration	25/07/2022	25/07/2022
							F08	Hospitalization or Prolonged Hospitalization	A010402	Migration	25/07/2022	25/07/2022
							F1905	Device Revision or Replacement	A010402	Migration	25/07/2022	25/07/2022
					CT145	Cardiovascular devices	E1716	Skin Discoloration	A010402	Migration	25/07/2022	25/07/2022
							E2403	No Clinical Signs, Symptoms or Conditions	A010402	Migration	25/07/2022	25/07/2022
							F08	Hospitalization or Prolonged Hospitalization	A010402	Migration	25/07/2022	25/07/2022
							F1905	Device Revision or Replacement	A010402	Migration	25/07/2022	25/07/2022
					CT1577	Wireless monitoring system transmitters	E1716	Skin Discoloration	A010402	Migration	25/07/2022	25/07/2022
							E2403	No Clinical Signs, Symptoms or Conditions	A010402	Migration	25/07/2022	25/07/2022
							F08	Hospitalization or Prolonged Hospitalization	A010402	Migration	25/07/2022	25/07/2022
							F1905	Device Revision or Replacement	A010402	Migration	25/07/2022	25/07/2022
					CT1684	Cardiac pacemakers and associated devices	E1716	Skin Discoloration	A010402	Migration	25/07/2022	25/07/2022
							E2403	No Clinical Signs, Symptoms or Conditions	A010402	Migration	25/07/2022	25/07/2022
							F08	Hospitalization or Prolonged Hospitalization	A010402	Migration	25/07/2022	25/07/2022
							F1905	Device Revision or Replacement	A010402	Migration	25/07/2022	25/07/2022

					CT2317	Transmitters/Receivers/Transceivers	E1716	Skin Discoloration	A010402	Migration	25/07/2022	25/07/2022
							E2403	No Clinical Signs, Symptoms or Conditions	A010402	Migration	25/07/2022	25/07/2022
							F08	Hospitalization or Prolonged Hospitalization	A010402	Migration	25/07/2022	25/07/2022
							F1905	Device Revision or Replacement	A010402	Migration	25/07/2022	25/07/2022
					CT233	Surgical	E1716	Skin Discoloration	A010402	Migration	25/07/2022	25/07/2022
							E2403	No Clinical Signs, Symptoms or Conditions	A010402	Migration	25/07/2022	25/07/2022
							F08	Hospitalization or Prolonged Hospitalization	A010402	Migration	25/07/2022	25/07/2022
							F1905	Device Revision or Replacement	A010402	Migration	25/07/2022	25/07/2022
					CT2406	Implantable	E1716	Skin Discoloration	A010402	Migration	25/07/2022	25/07/2022
							E2403	No Clinical Signs, Symptoms or Conditions	A010402	Migration	25/07/2022	25/07/2022
							F08	Hospitalization or Prolonged Hospitalization	A010402	Migration	25/07/2022	25/07/2022
							F1905	Device Revision or Replacement	A010402	Migration	25/07/2022	25/07/2022
					CT2407	Active implantable	E1716	Skin Discoloration	A010402	Migration	25/07/2022	25/07/2022
							E2403	No Clinical Signs, Symptoms or Conditions	A010402	Migration	25/07/2022	25/07/2022
							F08	Hospitalization or Prolonged Hospitalization	A010402	Migration	25/07/2022	25/07/2022
							F1905	Device Revision or Replacement	A010402	Migration	25/07/2022	25/07/2022
					CT255	Transducers and associated devices	E1716	Skin Discoloration	A010402	Migration	25/07/2022	25/07/2022
							E2403	No Clinical Signs, Symptoms or Conditions	A010402	Migration	25/07/2022	25/07/2022
							F08	Hospitalization or Prolonged Hospitalization	A010402	Migration	25/07/2022	25/07/2022
							F1905	Device Revision or Replacement	A010402	Migration	25/07/2022	25/07/2022
					CT2726	Physiological monitoring/measuring devices	E1716	Skin Discoloration	A010402	Migration	25/07/2022	25/07/2022
							E2403	No Clinical Signs, Symptoms or Conditions	A010402	Migration	25/07/2022	25/07/2022
							F08	Hospitalization or Prolonged Hospitalization	A010402	Migration	25/07/2022	25/07/2022
							F1905	Device Revision or Replacement	A010402	Migration	25/07/2022	25/07/2022
					CT2727	Active implant devices/ancillaries	E1716	Skin Discoloration	A010402	Migration	25/07/2022	25/07/2022
							E2403	No Clinical Signs, Symptoms or Conditions	A010402	Migration	25/07/2022	25/07/2022
							F08	Hospitalization or Prolonged Hospitalization	A010402	Migration	25/07/2022	25/07/2022
							F1905	Device Revision or Replacement	A010402	Migration	25/07/2022	25/07/2022
					CT2750	Implantable cardiac electrical stimulators and associated devices	E1716	Skin Discoloration	A010402	Migration	25/07/2022	25/07/2022
							E2403	No Clinical Signs, Symptoms or Conditions	A010402	Migration	25/07/2022	25/07/2022
							F08	Hospitalization or Prolonged Hospitalization	A010402	Migration	25/07/2022	25/07/2022
							F1905	Device Revision or Replacement	A010402	Migration	25/07/2022	25/07/2022
					CT2885	Physiological monitoring/measuring system transmitters/receivers/transceivers	E1716	Skin Discoloration	A010402	Migration	25/07/2022	25/07/2022
							E2403	No Clinical Signs, Symptoms or Conditions	A010402	Migration	25/07/2022	25/07/2022
							F08	Hospitalization or Prolonged Hospitalization	A010402	Migration	25/07/2022	25/07/2022
							F1905	Device Revision or Replacement	A010402	Migration	25/07/2022	25/07/2022
					CT2985	Active	E1716	Skin Discoloration	A010402	Migration	25/07/2022	25/07/2022
							E2403	No Clinical Signs, Symptoms or Conditions	A010402	Migration	25/07/2022	25/07/2022
							F08	Hospitalization or Prolonged Hospitalization	A010402	Migration	25/07/2022	25/07/2022
							F1905	Device Revision or Replacement	A010402	Migration	25/07/2022	25/07/2022
					CT3001	Implantable metal	E1716	Skin Discoloration	A010402	Migration	25/07/2022	25/07/2022
							E2403	No Clinical Signs, Symptoms or Conditions	A010402	Migration	25/07/2022	25/07/2022
							F08	Hospitalization or Prolonged Hospitalization	A010402	Migration	25/07/2022	25/07/2022
							F1905	Device Revision or Replacement	A010402	Migration	25/07/2022	25/07/2022
					CT3002	Implantable synthetic polymer	E1716	Skin Discoloration	A010402	Migration	25/07/2022	25/07/2022
							E2403	No Clinical Signs, Symptoms or Conditions	A010402	Migration	25/07/2022	25/07/2022
							F08	Hospitalization or Prolonged Hospitalization	A010402	Migration	25/07/2022	25/07/2022
							F1905	Device Revision or Replacement	A010402	Migration	25/07/2022	25/07/2022
					CT3006	Implantable inorganic material	E1716	Skin Discoloration	A010402	Migration	25/07/2022	25/07/2022
							E2403	No Clinical Signs, Symptoms or Conditions	A010402	Migration	25/07/2022	25/07/2022
							F08	Hospitalization or Prolonged Hospitalization	A010402	Migration	25/07/2022	25/07/2022
							F1905	Device Revision or Replacement	A010402	Migration	25/07/2022	25/07/2022
					CT314	Electrically powered	E1716	Skin Discoloration	A010402	Migration	25/07/2022	25/07/2022
							E2403	No Clinical Signs, Symptoms or Conditions	A010402	Migration	25/07/2022	25/07/2022
							F08	Hospitalization or Prolonged Hospitalization	A010402	Migration	25/07/2022	25/07/2022
							F1905	Device Revision or Replacement	A010402	Migration	25/07/2022	25/07/2022
					CT321	Long-term surgical invasive	E1716	Skin Discoloration	A010402	Migration	25/07/2022	25/07/2022
							E2403	No Clinical Signs, Symptoms or Conditions	A010402	Migration	25/07/2022	25/07/2022
							F08	Hospitalization or Prolonged Hospitalization	A010402	Migration	25/07/2022	25/07/2022
							F1905	Device Revision or Replacement	A010402	Migration	25/07/2022	25/07/2022
					CT384	Sensors and associated devices	E1716	Skin Discoloration	A010402	Migration	25/07/2022	25/07/2022
							E2403	No Clinical Signs, Symptoms or Conditions	A010402	Migration	25/07/2022	25/07/2022
							F08	Hospitalization or Prolonged Hospitalization	A010402	Migration	25/07/2022	25/07/2022
							F1905	Device Revision or Replacement	A010402	Migration	25/07/2022	25/07/2022
					CT449	Generators and associated devices	E1716	Skin Discoloration	A010402	Migration	25/07/2022	25/07/2022
							E2403	No Clinical Signs, Symptoms or Conditions	A010402	Migration	25/07/2022	25/07/2022
							F08	Hospitalization or Prolonged Hospitalization	A010402	Migration	25/07/2022	25/07/2022
							F1905	Device Revision or Replacement	A010402	Migration	25/07/2022	25/07/2022
					CT981	Single use	E1716	Skin Discoloration	A010402	Migration	25/07/2022	25/07/2022
							E2403	No Clinical Signs, Symptoms or Conditions	A010402	Migration	25/07/2022	25/07/2022
							F08	Hospitalization or Prolonged Hospitalization	A010402	Migration	25/07/2022	25/07/2022
							F1905	Device Revision or Replacement	A010402	Migration	25/07/2022	25/07/2022
					CT983	Surgical invasive	E1716	Skin Discoloration	A010402	Migration	25/07/2022	25/07/2022
							E2403	No Clinical Signs, Symptoms or Conditions	A010402	Migration	25/07/2022	25/07/2022
							F08	Hospitalization or Prolonged Hospitalization	A010402	Migration	25/07/2022	25/07/2022
							F1905	Device Revision or Replacement	A010402	Migration	25/07/2022	25/07/2022
					CT994	Sensors	E1716	Skin Discoloration	A010402	Migration	25/07/2022	25/07/2022
							E2403	No Clinical Signs, Symptoms or Conditions	A010402	Migration	25/07/2022	25/07/2022
							F08	Hospitalization or Prolonged Hospitalization	A010402	Migration	25/07/2022	25/07/2022
							F1905	Device Revision or Replacement	A010402	Migration	25/07/2022	25/07/2022
7463	2022	All other reportable incidents	36241	Pacemaker/defibrillator lead	CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A072201	High Impedance		
							F26	No Health Consequences or Impact	A072201	High Impedance		
7464	2022	All other reportable incidents	17882	Physiologic-monitoring defibrillation system	CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A071301	Failure to Deliver Shock/Stimulation		
							F26	No Health Consequences or Impact	A071301	Failure to Deliver Shock/Stimulation		
7465	2022	All other reportable incidents	36241	Pacemaker/defibrillator lead	CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070101	Failure to Capture		
							F26	No Health Consequences or Impact	A070101	Failure to Capture		
7466	2022	All other reportable incidents	35852	Single-chamber implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E2104	Electric Shock	A070909	Oversensing	04/08/2022	04/08/2022
									A071302	Inappropriate/Inadequate Shock/Stimulation	04/08/2022	04/08/2022

									A090801	Signal Artifact/Noise	04/08/2022	04/08/2022
							E2403	No Clinical Signs, Symptoms or Conditions	A070909	Oversensing	04/08/2022	04/08/2022
									A071302	Inappropriate/Inadequate Shock/Stimulation	04/08/2022	04/08/2022
									A090801	Signal Artifact/Noise	04/08/2022	04/08/2022
							F08	Hospitalization or Prolonged Hospitalization	A070909	Oversensing	04/08/2022	04/08/2022
									A071302	Inappropriate/Inadequate Shock/Stimulation	04/08/2022	04/08/2022
									A090801	Signal Artifact/Noise	04/08/2022	04/08/2022
							F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A070909	Oversensing	04/08/2022	04/08/2022
									A071302	Inappropriate/Inadequate Shock/Stimulation	04/08/2022	04/08/2022
									A090801	Signal Artifact/Noise	04/08/2022	04/08/2022
							F1905	Device Revision or Replacement	A070909	Oversensing	04/08/2022	04/08/2022
									A071302	Inappropriate/Inadequate Shock/Stimulation	04/08/2022	04/08/2022
									A090801	Signal Artifact/Noise	04/08/2022	04/08/2022
7467	2022	All other reportable incidents	36241	Pacemaker/defibrillator lead	CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070102	High Capture Threshold		
									A072201	High Impedance		
									A090801	Signal Artifact/Noise		
							F26	No Health Consequences or Impact	A070102	High Capture Threshold		
									A072201	High Impedance		
									A090801	Signal Artifact/Noise		
7468	2021	All other reportable incidents	35852	Single-chamber implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E1713	Pocket Erosion	A24	No Apparent Device Problem	14/12/2021	14/12/2021
							E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	14/12/2021	14/12/2021
							F08	Hospitalization or Prolonged Hospitalization	A24	No Apparent Device Problem	14/12/2021	14/12/2021
							F1905	Device Revision or Replacement	A24	No Apparent Device Problem	14/12/2021	14/12/2021
7469	2022	All other reportable incidents	35223	Endocardial/interventricular septal pacing lead	CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070102	High Capture Threshold		
									A072201	High Impedance		
							F26	No Health Consequences or Impact	A070102	High Capture Threshold		
									A072201	High Impedance		
7470	2022	All other reportable incidents	35223	Endocardial/interventricular septal pacing lead	CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070909	Oversensing	07/09/2022	07/09/2022
									A0712	Pacing Problem	07/09/2022	07/09/2022
									A090801	Signal Artifact/Noise	07/09/2022	07/09/2022
							F12	Serious Injury/ Illness/ Impairment	A070909	Oversensing	07/09/2022	07/09/2022
									A0712	Pacing Problem	07/09/2022	07/09/2022
									A090801	Signal Artifact/Noise	07/09/2022	07/09/2022
							F1905	Device Revision or Replacement	A070909	Oversensing	07/09/2022	07/09/2022
									A0712	Pacing Problem	07/09/2022	07/09/2022
									A090801	Signal Artifact/Noise	07/09/2022	07/09/2022
7471	2022	All other reportable incidents	47265	Dual-chamber implantable pacemaker, rate-responsive	CT141	Electrodes/Leads and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A160503	Inappropriate or Unexpected Reset	01/08/2022	01/08/2022
							F08	Hospitalization or Prolonged Hospitalization	A160503	Inappropriate or Unexpected Reset	01/08/2022	01/08/2022
							F1905	Device Revision or Replacement	A160503	Inappropriate or Unexpected Reset	01/08/2022	01/08/2022
					CT145	Cardiovascular devices	E2403	No Clinical Signs, Symptoms or Conditions	A160503	Inappropriate or Unexpected Reset	01/08/2022	01/08/2022
							F08	Hospitalization or Prolonged Hospitalization	A160503	Inappropriate or Unexpected Reset	01/08/2022	01/08/2022
							F1905	Device Revision or Replacement	A160503	Inappropriate or Unexpected Reset	01/08/2022	01/08/2022
					CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A160503	Inappropriate or Unexpected Reset	01/08/2022	01/08/2022
							F08	Hospitalization or Prolonged Hospitalization	A160503	Inappropriate or Unexpected Reset	01/08/2022	01/08/2022
							F1905	Device Revision or Replacement	A160503	Inappropriate or Unexpected Reset	01/08/2022	01/08/2022
					CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A160503	Inappropriate or Unexpected Reset	01/08/2022	01/08/2022
							F08	Hospitalization or Prolonged Hospitalization	A160503	Inappropriate or Unexpected Reset	01/08/2022	01/08/2022
							F1905	Device Revision or Replacement	A160503	Inappropriate or Unexpected Reset	01/08/2022	01/08/2022
					CT233	Surgical	E2403	No Clinical Signs, Symptoms or Conditions	A160503	Inappropriate or Unexpected Reset	01/08/2022	01/08/2022
							F08	Hospitalization or Prolonged Hospitalization	A160503	Inappropriate or Unexpected Reset	01/08/2022	01/08/2022
							F1905	Device Revision or Replacement	A160503	Inappropriate or Unexpected Reset	01/08/2022	01/08/2022
					CT2406	Implantable	E2403	No Clinical Signs, Symptoms or Conditions	A160503	Inappropriate or Unexpected Reset	01/08/2022	01/08/2022
							F08	Hospitalization or Prolonged Hospitalization	A160503	Inappropriate or Unexpected Reset	01/08/2022	01/08/2022
							F1905	Device Revision or Replacement	A160503	Inappropriate or Unexpected Reset	01/08/2022	01/08/2022
					CT2407	Active implantable	E2403	No Clinical Signs, Symptoms or Conditions	A160503	Inappropriate or Unexpected Reset	01/08/2022	01/08/2022
							F08	Hospitalization or Prolonged Hospitalization	A160503	Inappropriate or Unexpected Reset	01/08/2022	01/08/2022
							F1905	Device Revision or Replacement	A160503	Inappropriate or Unexpected Reset	01/08/2022	01/08/2022
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A160503	Inappropriate or Unexpected Reset	01/08/2022	01/08/2022
							F08	Hospitalization or Prolonged Hospitalization	A160503	Inappropriate or Unexpected Reset	01/08/2022	01/08/2022
							F1905	Device Revision or Replacement	A160503	Inappropriate or Unexpected Reset	01/08/2022	01/08/2022
					CT2727	Active implant devices/ancillaries	E2403	No Clinical Signs, Symptoms or Conditions	A160503	Inappropriate or Unexpected Reset	01/08/2022	01/08/2022
							F08	Hospitalization or Prolonged Hospitalization	A160503	Inappropriate or Unexpected Reset	01/08/2022	01/08/2022
							F1905	Device Revision or Replacement	A160503	Inappropriate or Unexpected Reset	01/08/2022	01/08/2022
					CT2750	Implantable cardiac electrical stimulators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A160503	Inappropriate or Unexpected Reset	01/08/2022	01/08/2022
							F08	Hospitalization or Prolonged Hospitalization	A160503	Inappropriate or Unexpected Reset	01/08/2022	01/08/2022
							F1905	Device Revision or Replacement	A160503	Inappropriate or Unexpected Reset	01/08/2022	01/08/2022
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A160503	Inappropriate or Unexpected Reset	01/08/2022	01/08/2022
							F08	Hospitalization or Prolonged Hospitalization	A160503	Inappropriate or Unexpected Reset	01/08/2022	01/08/2022
							F1905	Device Revision or Replacement	A160503	Inappropriate or Unexpected Reset	01/08/2022	01/08/2022
					CT2985	Active	E2403	No Clinical Signs, Symptoms or Conditions	A160503	Inappropriate or Unexpected Reset	01/08/2022	01/08/2022
							F08	Hospitalization or Prolonged Hospitalization	A160503	Inappropriate or Unexpected Reset	01/08/2022	01/08/2022
							F1905	Device Revision or Replacement	A160503	Inappropriate or Unexpected Reset	01/08/2022	01/08/2022
					CT3001	Implantable metal	E2403	No Clinical Signs, Symptoms or Conditions	A160503	Inappropriate or Unexpected Reset	01/08/2022	01/08/2022
							F08	Hospitalization or Prolonged Hospitalization	A160503	Inappropriate or Unexpected Reset	01/08/2022	01/08/2022
							F1905	Device Revision or Replacement	A160503	Inappropriate or Unexpected Reset	01/08/2022	01/08/2022
					CT3002	Implantable synthetic polymer	E2403	No Clinical Signs, Symptoms or Conditions	A160503	Inappropriate or Unexpected Reset	01/08/2022	01/08/2022
							F08	Hospitalization or Prolonged Hospitalization	A160503	Inappropriate or Unexpected Reset	01/08/2022	01/08/2022
							F1905	Device Revision or Replacement	A160503	Inappropriate or Unexpected Reset	01/08/2022	01/08/2022
					CT3006	Implantable inorganic material	E2403	No Clinical Signs, Symptoms or Conditions	A160503	Inappropriate or Unexpected Reset	01/08/2022	01/08/2022
							F08	Hospitalization or Prolonged Hospitalization	A160503	Inappropriate or Unexpected Reset	01/08/2022	01/08/2022
							F1905	Device Revision or Replacement	A160503	Inappropriate or Unexpected Reset	01/08/2022	01/08/2022
					CT314	Electrically powered	E2403	No Clinical Signs, Symptoms or Conditions	A160503	Inappropriate or Unexpected Reset	01/08/2022	01/08/2022
							F08	Hospitalization or Prolonged Hospitalization	A160503	Inappropriate or Unexpected Reset	01/08/2022	01/08/2022
							F1905	Device Revision or Replacement	A160503	Inappropriate or Unexpected Reset	01/08/2022	01/08/2022
					CT321	Long-term surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A160503	Inappropriate or Unexpected Reset	01/08/2022	01/08/2022
							F08	Hospitalization or Prolonged Hospitalization	A160503	Inappropriate or Unexpected Reset	01/08/2022	01/08/2022
							F1905	Device Revision or Replacement	A160503	Inappropriate or Unexpected Reset	01/08/2022	01/08/2022
					CT449	Generators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A160503	Inappropriate or Unexpected Reset	01/08/2022	01/08/2022

							F08	Hospitalization or Prolonged Hospitalization	A160503	Inappropriate or Unexpected Reset	01/08/2022	01/08/2022
							F1905	Device Revision or Replacement	A160503	Inappropriate or Unexpected Reset	01/08/2022	01/08/2022
					CT981	Single use	E2403	No Clinical Signs, Symptoms or Conditions	A160503	Inappropriate or Unexpected Reset	01/08/2022	01/08/2022
							F08	Hospitalization or Prolonged Hospitalization	A160503	Inappropriate or Unexpected Reset	01/08/2022	01/08/2022
							F1905	Device Revision or Replacement	A160503	Inappropriate or Unexpected Reset	01/08/2022	01/08/2022
					CT983	Surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A160503	Inappropriate or Unexpected Reset	01/08/2022	01/08/2022
							F08	Hospitalization or Prolonged Hospitalization	A160503	Inappropriate or Unexpected Reset	01/08/2022	01/08/2022
							F1905	Device Revision or Replacement	A160503	Inappropriate or Unexpected Reset	01/08/2022	01/08/2022
7472	2022	All other reportable incidents	17882	Physiologic-monitoring defibrillation system	CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A07	Electrical /Electronic Property Problem		
							F27	Problem Identified During Non-Clinical Procedure	A07	Electrical /Electronic Property Problem		
7473	2022	All other reportable incidents	17882	Physiologic-monitoring defibrillation system	CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070908	Failure to Sense		
							F26	No Health Consequences or Impact	A070908	Failure to Sense		
7474	2022	All other reportable incidents	47265	Dual-chamber implantable pacemaker, rate-responsive	CT141	Electrodes/Leads and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	22/02/2022	22/02/2022
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	22/02/2022	22/02/2022
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	22/02/2022	22/02/2022
					CT145	Cardiovascular devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	22/02/2022	22/02/2022
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	22/02/2022	22/02/2022
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	22/02/2022	22/02/2022
					CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	22/02/2022	22/02/2022
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	22/02/2022	22/02/2022
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	22/02/2022	22/02/2022
					CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	22/02/2022	22/02/2022
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	22/02/2022	22/02/2022
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	22/02/2022	22/02/2022
					CT233	Surgical	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	22/02/2022	22/02/2022
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	22/02/2022	22/02/2022
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	22/02/2022	22/02/2022
					CT2406	Implantable	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	22/02/2022	22/02/2022
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	22/02/2022	22/02/2022
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	22/02/2022	22/02/2022
					CT2407	Active implantable	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	22/02/2022	22/02/2022
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	22/02/2022	22/02/2022
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	22/02/2022	22/02/2022
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	22/02/2022	22/02/2022
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	22/02/2022	22/02/2022
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	22/02/2022	22/02/2022
					CT2727	Active implant devices/ancillaries	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	22/02/2022	22/02/2022
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	22/02/2022	22/02/2022
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	22/02/2022	22/02/2022
					CT2750	Implantable cardiac electrical stimulators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	22/02/2022	22/02/2022
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	22/02/2022	22/02/2022
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	22/02/2022	22/02/2022
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	22/02/2022	22/02/2022
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	22/02/2022	22/02/2022
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	22/02/2022	22/02/2022
					CT2985	Active	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	22/02/2022	22/02/2022
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	22/02/2022	22/02/2022
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	22/02/2022	22/02/2022
					CT3001	Implantable metal	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	22/02/2022	22/02/2022
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	22/02/2022	22/02/2022
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	22/02/2022	22/02/2022
					CT3002	Implantable synthetic polymer	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	22/02/2022	22/02/2022
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	22/02/2022	22/02/2022
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	22/02/2022	22/02/2022
					CT3006	Implantable inorganic material	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	22/02/2022	22/02/2022
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	22/02/2022	22/02/2022
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	22/02/2022	22/02/2022
					CT314	Electrically powered	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	22/02/2022	22/02/2022
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	22/02/2022	22/02/2022
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	22/02/2022	22/02/2022
					CT321	Long-term surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	22/02/2022	22/02/2022
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	22/02/2022	22/02/2022
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	22/02/2022	22/02/2022
					CT449	Generators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	22/02/2022	22/02/2022
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	22/02/2022	22/02/2022
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	22/02/2022	22/02/2022
					CT981	Single use	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	22/02/2022	22/02/2022
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	22/02/2022	22/02/2022
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	22/02/2022	22/02/2022
					CT983	Surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	22/02/2022	22/02/2022
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	22/02/2022	22/02/2022
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	22/02/2022	22/02/2022
7475	2022	All other reportable incidents	17882	Physiologic-monitoring defibrillation system	CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A0801	Failure to Calibrate		
							F27	Problem Identified During Non-Clinical Procedure	A0801	Failure to Calibrate		
7476	2022	Death	47265	Dual-chamber implantable pacemaker, rate-responsive	CT141	Electrodes/Leads and associated devices	E2401	Insufficient Information	A13	Communication or Transmission Problem		
							F02	Death	A13	Communication or Transmission Problem		
					CT145	Cardiovascular devices	E2401	Insufficient Information	A13	Communication or Transmission Problem		
							F02	Death	A13	Communication or Transmission Problem		
					CT1684	Cardiac pacemakers and associated devices	E2401	Insufficient Information	A13	Communication or Transmission Problem		
							F02	Death	A13	Communication or Transmission Problem		
					CT1685	Cardiac pacemakers	E2401	Insufficient Information	A13	Communication or Transmission Problem		
							F02	Death	A13	Communication or Transmission Problem		
					CT233	Surgical	E2401	Insufficient Information	A13	Communication or Transmission Problem		
							F02	Death	A13	Communication or Transmission Problem		
					CT2406	Implantable	E2401	Insufficient Information	A13	Communication or Transmission Problem		
							F02	Death	A13	Communication or Transmission Problem		
					CT2407	Active implantable	E2401	Insufficient Information	A13	Communication or Transmission Problem		
							F02	Death	A13	Communication or Transmission Problem		

					CT257	Implantable cardiac pacemakers	E2401	Insufficient Information	A13	Communication or Transmission Problem		
							F02	Death	A13	Communication or Transmission Problem		
					CT2727	Active implant devices/ancillaries	E2401	Insufficient Information	A13	Communication or Transmission Problem		
							F02	Death	A13	Communication or Transmission Problem		
					CT2750	Implantable cardiac electrical stimulators and associated devices	E2401	Insufficient Information	A13	Communication or Transmission Problem		
							F02	Death	A13	Communication or Transmission Problem		
					CT2866	Implantable cardiac electrical stimulators	E2401	Insufficient Information	A13	Communication or Transmission Problem		
							F02	Death	A13	Communication or Transmission Problem		
					CT2985	Active	E2401	Insufficient Information	A13	Communication or Transmission Problem		
							F02	Death	A13	Communication or Transmission Problem		
					CT3001	Implantable metal	E2401	Insufficient Information	A13	Communication or Transmission Problem		
							F02	Death	A13	Communication or Transmission Problem		
					CT3002	Implantable synthetic polymer	E2401	Insufficient Information	A13	Communication or Transmission Problem		
							F02	Death	A13	Communication or Transmission Problem		
					CT3006	Implantable inorganic material	E2401	Insufficient Information	A13	Communication or Transmission Problem		
							F02	Death	A13	Communication or Transmission Problem		
					CT314	Electrically powered	E2401	Insufficient Information	A13	Communication or Transmission Problem		
							F02	Death	A13	Communication or Transmission Problem		
					CT321	Long-term surgical invasive	E2401	Insufficient Information	A13	Communication or Transmission Problem		
							F02	Death	A13	Communication or Transmission Problem		
					CT449	Generators and associated devices	E2401	Insufficient Information	A13	Communication or Transmission Problem		
							F02	Death	A13	Communication or Transmission Problem		
					CT981	Single use	E2401	Insufficient Information	A13	Communication or Transmission Problem		
							F02	Death	A13	Communication or Transmission Problem		
					CT983	Surgical invasive	E2401	Insufficient Information	A13	Communication or Transmission Problem		
							F02	Death	A13	Communication or Transmission Problem		
7477	2022	All other reportable incidents	47265	Dual-chamber implantable pacemaker, rate-responsive	CT141	Electrodes/Leads and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem	09/10/2022	09/10/2022
									A070504	Premature Discharge of Battery	09/10/2022	09/10/2022
							F08	Hospitalization or Prolonged Hospitalization	A0705	Battery Problem	09/10/2022	09/10/2022
									A070504	Premature Discharge of Battery	09/10/2022	09/10/2022
							F1905	Device Revision or Replacement	A0705	Battery Problem	09/10/2022	09/10/2022
									A070504	Premature Discharge of Battery	09/10/2022	09/10/2022
					CT145	Cardiovascular devices	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem	09/10/2022	09/10/2022
									A070504	Premature Discharge of Battery	09/10/2022	09/10/2022
							F08	Hospitalization or Prolonged Hospitalization	A0705	Battery Problem	09/10/2022	09/10/2022
									A070504	Premature Discharge of Battery	09/10/2022	09/10/2022
							F1905	Device Revision or Replacement	A0705	Battery Problem	09/10/2022	09/10/2022
									A070504	Premature Discharge of Battery	09/10/2022	09/10/2022
					CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem	09/10/2022	09/10/2022
									A070504	Premature Discharge of Battery	09/10/2022	09/10/2022
							F08	Hospitalization or Prolonged Hospitalization	A0705	Battery Problem	09/10/2022	09/10/2022
									A070504	Premature Discharge of Battery	09/10/2022	09/10/2022
							F1905	Device Revision or Replacement	A0705	Battery Problem	09/10/2022	09/10/2022
									A070504	Premature Discharge of Battery	09/10/2022	09/10/2022
					CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem	09/10/2022	09/10/2022
									A070504	Premature Discharge of Battery	09/10/2022	09/10/2022
							F08	Hospitalization or Prolonged Hospitalization	A0705	Battery Problem	09/10/2022	09/10/2022
									A070504	Premature Discharge of Battery	09/10/2022	09/10/2022
							F1905	Device Revision or Replacement	A0705	Battery Problem	09/10/2022	09/10/2022
									A070504	Premature Discharge of Battery	09/10/2022	09/10/2022
					CT233	Surgical	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem	09/10/2022	09/10/2022
									A070504	Premature Discharge of Battery	09/10/2022	09/10/2022
							F08	Hospitalization or Prolonged Hospitalization	A0705	Battery Problem	09/10/2022	09/10/2022
									A070504	Premature Discharge of Battery	09/10/2022	09/10/2022
							F1905	Device Revision or Replacement	A0705	Battery Problem	09/10/2022	09/10/2022
									A070504	Premature Discharge of Battery	09/10/2022	09/10/2022
					CT2406	Implantable	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem	09/10/2022	09/10/2022
									A070504	Premature Discharge of Battery	09/10/2022	09/10/2022
							F08	Hospitalization or Prolonged Hospitalization	A0705	Battery Problem	09/10/2022	09/10/2022
									A070504	Premature Discharge of Battery	09/10/2022	09/10/2022
							F1905	Device Revision or Replacement	A0705	Battery Problem	09/10/2022	09/10/2022
									A070504	Premature Discharge of Battery	09/10/2022	09/10/2022
					CT2407	Active implantable	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem	09/10/2022	09/10/2022
									A070504	Premature Discharge of Battery	09/10/2022	09/10/2022
							F08	Hospitalization or Prolonged Hospitalization	A0705	Battery Problem	09/10/2022	09/10/2022
									A070504	Premature Discharge of Battery	09/10/2022	09/10/2022
							F1905	Device Revision or Replacement	A0705	Battery Problem	09/10/2022	09/10/2022
									A070504	Premature Discharge of Battery	09/10/2022	09/10/2022
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem	09/10/2022	09/10/2022
									A070504	Premature Discharge of Battery	09/10/2022	09/10/2022
							F08	Hospitalization or Prolonged Hospitalization	A0705	Battery Problem	09/10/2022	09/10/2022
									A070504	Premature Discharge of Battery	09/10/2022	09/10/2022
							F1905	Device Revision or Replacement	A0705	Battery Problem	09/10/2022	09/10/2022
									A070504	Premature Discharge of Battery	09/10/2022	09/10/2022
					CT2727	Active implant devices/ancillaries	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem	09/10/2022	09/10/2022
									A070504	Premature Discharge of Battery	09/10/2022	09/10/2022
							F08	Hospitalization or Prolonged Hospitalization	A0705	Battery Problem	09/10/2022	09/10/2022
									A070504	Premature Discharge of Battery	09/10/2022	09/10/2022
							F1905	Device Revision or Replacement	A0705	Battery Problem	09/10/2022	09/10/2022
									A070504	Premature Discharge of Battery	09/10/2022	09/10/2022
					CT2750	Implantable cardiac electrical stimulators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem	09/10/2022	09/10/2022
									A070504	Premature Discharge of Battery	09/10/2022	09/10/2022
							F08	Hospitalization or Prolonged Hospitalization	A0705	Battery Problem	09/10/2022	09/10/2022
									A070504	Premature Discharge of Battery	09/10/2022	09/10/2022
							F1905	Device Revision or Replacement	A0705	Battery Problem	09/10/2022	09/10/2022
									A070504	Premature Discharge of Battery	09/10/2022	09/10/2022
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem	09/10/2022	09/10/2022
									A070504	Premature Discharge of Battery	09/10/2022	09/10/2022

							F08	Hospitalization or Prolonged Hospitalization	A0705	Battery Problem	09/10/2022	09/10/2022
									A070504	Premature Discharge of Battery	09/10/2022	09/10/2022
							F1905	Device Revision or Replacement	A0705	Battery Problem	09/10/2022	09/10/2022
									A070504	Premature Discharge of Battery	09/10/2022	09/10/2022
					CT2985	Active	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem	09/10/2022	09/10/2022
									A070504	Premature Discharge of Battery	09/10/2022	09/10/2022
							F08	Hospitalization or Prolonged Hospitalization	A0705	Battery Problem	09/10/2022	09/10/2022
									A070504	Premature Discharge of Battery	09/10/2022	09/10/2022
							F1905	Device Revision or Replacement	A0705	Battery Problem	09/10/2022	09/10/2022
									A070504	Premature Discharge of Battery	09/10/2022	09/10/2022
					CT3001	Implantable metal	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem	09/10/2022	09/10/2022
									A070504	Premature Discharge of Battery	09/10/2022	09/10/2022
							F08	Hospitalization or Prolonged Hospitalization	A0705	Battery Problem	09/10/2022	09/10/2022
									A070504	Premature Discharge of Battery	09/10/2022	09/10/2022
							F1905	Device Revision or Replacement	A0705	Battery Problem	09/10/2022	09/10/2022
									A070504	Premature Discharge of Battery	09/10/2022	09/10/2022
					CT3002	Implantable synthetic polymer	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem	09/10/2022	09/10/2022
									A070504	Premature Discharge of Battery	09/10/2022	09/10/2022
							F08	Hospitalization or Prolonged Hospitalization	A0705	Battery Problem	09/10/2022	09/10/2022
									A070504	Premature Discharge of Battery	09/10/2022	09/10/2022
							F1905	Device Revision or Replacement	A0705	Battery Problem	09/10/2022	09/10/2022
									A070504	Premature Discharge of Battery	09/10/2022	09/10/2022
					CT3006	Implantable inorganic material	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem	09/10/2022	09/10/2022
									A070504	Premature Discharge of Battery	09/10/2022	09/10/2022
							F08	Hospitalization or Prolonged Hospitalization	A0705	Battery Problem	09/10/2022	09/10/2022
									A070504	Premature Discharge of Battery	09/10/2022	09/10/2022
							F1905	Device Revision or Replacement	A0705	Battery Problem	09/10/2022	09/10/2022
									A070504	Premature Discharge of Battery	09/10/2022	09/10/2022
					CT314	Electrically powered	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem	09/10/2022	09/10/2022
									A070504	Premature Discharge of Battery	09/10/2022	09/10/2022
							F08	Hospitalization or Prolonged Hospitalization	A0705	Battery Problem	09/10/2022	09/10/2022
									A070504	Premature Discharge of Battery	09/10/2022	09/10/2022
							F1905	Device Revision or Replacement	A0705	Battery Problem	09/10/2022	09/10/2022
									A070504	Premature Discharge of Battery	09/10/2022	09/10/2022
					CT321	Long-term surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem	09/10/2022	09/10/2022
									A070504	Premature Discharge of Battery	09/10/2022	09/10/2022
							F08	Hospitalization or Prolonged Hospitalization	A0705	Battery Problem	09/10/2022	09/10/2022
									A070504	Premature Discharge of Battery	09/10/2022	09/10/2022
							F1905	Device Revision or Replacement	A0705	Battery Problem	09/10/2022	09/10/2022
									A070504	Premature Discharge of Battery	09/10/2022	09/10/2022
					CT449	Generators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem	09/10/2022	09/10/2022
									A070504	Premature Discharge of Battery	09/10/2022	09/10/2022
							F08	Hospitalization or Prolonged Hospitalization	A0705	Battery Problem	09/10/2022	09/10/2022
									A070504	Premature Discharge of Battery	09/10/2022	09/10/2022
							F1905	Device Revision or Replacement	A0705	Battery Problem	09/10/2022	09/10/2022
									A070504	Premature Discharge of Battery	09/10/2022	09/10/2022
					CT981	Single use	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem	09/10/2022	09/10/2022
									A070504	Premature Discharge of Battery	09/10/2022	09/10/2022
							F08	Hospitalization or Prolonged Hospitalization	A0705	Battery Problem	09/10/2022	09/10/2022
									A070504	Premature Discharge of Battery	09/10/2022	09/10/2022
							F1905	Device Revision or Replacement	A0705	Battery Problem	09/10/2022	09/10/2022
									A070504	Premature Discharge of Battery	09/10/2022	09/10/2022
					CT983	Surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem	09/10/2022	09/10/2022
									A070504	Premature Discharge of Battery	09/10/2022	09/10/2022
							F08	Hospitalization or Prolonged Hospitalization	A0705	Battery Problem	09/10/2022	09/10/2022
									A070504	Premature Discharge of Battery	09/10/2022	09/10/2022
							F1905	Device Revision or Replacement	A0705	Battery Problem	09/10/2022	09/10/2022
									A070504	Premature Discharge of Battery	09/10/2022	09/10/2022
7478	2022	All other reportable incidents	47265	Dual-chamber implantable pacemaker, rate-responsive	CT141	Electrodes/Leads and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT145	Cardiovascular devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT233	Surgical	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT2406	Implantable	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT2407	Active implantable	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT2727	Active implant devices/ancillaries	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT2750	Implantable cardiac electrical stimulators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT2985	Active	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT3001	Implantable metal	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT3002	Implantable synthetic polymer	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT3006	Implantable inorganic material	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		

					CT314	Electrically powered	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery			
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery			
					CT321	Long-term surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery			
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery			
					CT449	Generators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery			
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery			
					CT981	Single use	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery			
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery			
					CT983	Surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery			
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery			
7479	2022	All other reportable incidents	47265	Dual-chamber implantable pacemaker, rate-responsive	CT141	Electrodes/Leads and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	01/09/2022	01/09/2022
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	01/09/2022	01/09/2022
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	01/09/2022	01/09/2022
					CT145	Cardiovascular devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	01/09/2022	01/09/2022
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	01/09/2022	01/09/2022
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	01/09/2022	01/09/2022
					CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	01/09/2022	01/09/2022
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	01/09/2022	01/09/2022
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	01/09/2022	01/09/2022
					CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	01/09/2022	01/09/2022
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	01/09/2022	01/09/2022
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	01/09/2022	01/09/2022
					CT233	Surgical	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	01/09/2022	01/09/2022
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	01/09/2022	01/09/2022
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	01/09/2022	01/09/2022
					CT2406	Implantable	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	01/09/2022	01/09/2022
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	01/09/2022	01/09/2022
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	01/09/2022	01/09/2022
					CT2407	Active implantable	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	01/09/2022	01/09/2022
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	01/09/2022	01/09/2022
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	01/09/2022	01/09/2022
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	01/09/2022	01/09/2022
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	01/09/2022	01/09/2022
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	01/09/2022	01/09/2022
					CT2727	Active implant devices/ancillaries	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	01/09/2022	01/09/2022
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	01/09/2022	01/09/2022
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	01/09/2022	01/09/2022
					CT2750	Implantable cardiac electrical stimulators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	01/09/2022	01/09/2022
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	01/09/2022	01/09/2022
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	01/09/2022	01/09/2022
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	01/09/2022	01/09/2022
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	01/09/2022	01/09/2022
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	01/09/2022	01/09/2022
					CT2985	Active	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	01/09/2022	01/09/2022
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	01/09/2022	01/09/2022
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	01/09/2022	01/09/2022
					CT3001	Implantable metal	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	01/09/2022	01/09/2022
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	01/09/2022	01/09/2022
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	01/09/2022	01/09/2022
					CT3002	Implantable synthetic polymer	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	01/09/2022	01/09/2022
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	01/09/2022	01/09/2022
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	01/09/2022	01/09/2022
					CT3006	Implantable inorganic material	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	01/09/2022	01/09/2022
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	01/09/2022	01/09/2022
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	01/09/2022	01/09/2022
					CT314	Electrically powered	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	01/09/2022	01/09/2022
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	01/09/2022	01/09/2022
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	01/09/2022	01/09/2022
					CT321	Long-term surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	01/09/2022	01/09/2022
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	01/09/2022	01/09/2022
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	01/09/2022	01/09/2022
					CT449	Generators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	01/09/2022	01/09/2022
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	01/09/2022	01/09/2022
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	01/09/2022	01/09/2022
					CT981	Single use	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	01/09/2022	01/09/2022
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	01/09/2022	01/09/2022
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	01/09/2022	01/09/2022
					CT983	Surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	01/09/2022	01/09/2022
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	01/09/2022	01/09/2022
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	01/09/2022	01/09/2022
7480	2022	All other reportable incidents	36241	Pacemaker/defibrillator lead	CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A051201	Device Dislodged or Dislocated	09/12/2021	09/12/2021
									A070102	High Capture Threshold	09/12/2021	09/12/2021
							F08	Hospitalization or Prolonged Hospitalization	A051201	Device Dislodged or Dislocated	09/12/2021	09/12/2021
									A070102	High Capture Threshold	09/12/2021	09/12/2021
							F1905	Device Revision or Replacement	A051201	Device Dislodged or Dislocated	09/12/2021	09/12/2021
									A070102	High Capture Threshold	09/12/2021	09/12/2021
							F2203	Imaging Required	A051201	Device Dislodged or Dislocated	09/12/2021	09/12/2021
									A070102	High Capture Threshold	09/12/2021	09/12/2021
7481	2022	All other reportable incidents	47270	Cardiac resynchronization therapy implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A070909	Oversensing		
							F26	No Health Consequences or Impact	A070909	Oversensing		
7482	2022	All other reportable incidents	17882	Physiologic-monitoring defibrillation system	CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A0710	Failure to Conduct		
							F27	Problem Identified During Non-Clinical Procedure	A0710	Failure to Conduct		
7483	2022	All other reportable incidents	47265	Dual-chamber implantable pacemaker, rate-responsive	CT141	Electrodes/Leads and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	07/02/2022	07/02/2022
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	07/02/2022	07/02/2022
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	07/02/2022	07/02/2022
					CT145	Cardiovascular devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	07/02/2022	07/02/2022
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	07/02/2022	07/02/2022
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	07/02/2022	07/02/2022

					CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	07/02/2022	07/02/2022
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	07/02/2022	07/02/2022
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	07/02/2022	07/02/2022
					CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	07/02/2022	07/02/2022
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	07/02/2022	07/02/2022
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	07/02/2022	07/02/2022
					CT233	Surgical	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	07/02/2022	07/02/2022
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	07/02/2022	07/02/2022
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	07/02/2022	07/02/2022
					CT2406	Implantable	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	07/02/2022	07/02/2022
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	07/02/2022	07/02/2022
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	07/02/2022	07/02/2022
					CT2407	Active implantable	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	07/02/2022	07/02/2022
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	07/02/2022	07/02/2022
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	07/02/2022	07/02/2022
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	07/02/2022	07/02/2022
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	07/02/2022	07/02/2022
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	07/02/2022	07/02/2022
					CT2727	Active implant devices/ancillaries	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	07/02/2022	07/02/2022
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	07/02/2022	07/02/2022
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	07/02/2022	07/02/2022
					CT2750	Implantable cardiac electrical stimulators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	07/02/2022	07/02/2022
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	07/02/2022	07/02/2022
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	07/02/2022	07/02/2022
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	07/02/2022	07/02/2022
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	07/02/2022	07/02/2022
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	07/02/2022	07/02/2022
					CT2985	Active	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	07/02/2022	07/02/2022
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	07/02/2022	07/02/2022
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	07/02/2022	07/02/2022
					CT3001	Implantable metal	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	07/02/2022	07/02/2022
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	07/02/2022	07/02/2022
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	07/02/2022	07/02/2022
					CT3002	Implantable synthetic polymer	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	07/02/2022	07/02/2022
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	07/02/2022	07/02/2022
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	07/02/2022	07/02/2022
					CT3006	Implantable inorganic material	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	07/02/2022	07/02/2022
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	07/02/2022	07/02/2022
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	07/02/2022	07/02/2022
					CT314	Electrically powered	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	07/02/2022	07/02/2022
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	07/02/2022	07/02/2022
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	07/02/2022	07/02/2022
					CT321	Long-term surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	07/02/2022	07/02/2022
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	07/02/2022	07/02/2022
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	07/02/2022	07/02/2022
					CT449	Generators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	07/02/2022	07/02/2022
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	07/02/2022	07/02/2022
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	07/02/2022	07/02/2022
					CT981	Single use	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	07/02/2022	07/02/2022
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	07/02/2022	07/02/2022
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	07/02/2022	07/02/2022
					CT983	Surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	07/02/2022	07/02/2022
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	07/02/2022	07/02/2022
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	07/02/2022	07/02/2022
7484	2022	All other reportable incidents	47265	Dual-chamber implantable pacemaker, rate-responsive	CT141	Electrodes/Leads and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A0908	Incorrect, Inadequate or Imprecise Result or Readings	05/08/2022	05/08/2022
							F08	Hospitalization or Prolonged Hospitalization	A0908	Incorrect, Inadequate or Imprecise Result or Readings	05/08/2022	05/08/2022
							F1903	Device Explantation	A0908	Incorrect, Inadequate or Imprecise Result or Readings	05/08/2022	05/08/2022
					CT145	Cardiovascular devices	E2403	No Clinical Signs, Symptoms or Conditions	A0908	Incorrect, Inadequate or Imprecise Result or Readings	05/08/2022	05/08/2022
							F08	Hospitalization or Prolonged Hospitalization	A0908	Incorrect, Inadequate or Imprecise Result or Readings	05/08/2022	05/08/2022
							F1903	Device Explantation	A0908	Incorrect, Inadequate or Imprecise Result or Readings	05/08/2022	05/08/2022
					CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A0908	Incorrect, Inadequate or Imprecise Result or Readings	05/08/2022	05/08/2022
							F08	Hospitalization or Prolonged Hospitalization	A0908	Incorrect, Inadequate or Imprecise Result or Readings	05/08/2022	05/08/2022
							F1903	Device Explantation	A0908	Incorrect, Inadequate or Imprecise Result or Readings	05/08/2022	05/08/2022
					CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A0908	Incorrect, Inadequate or Imprecise Result or Readings	05/08/2022	05/08/2022
							F08	Hospitalization or Prolonged Hospitalization	A0908	Incorrect, Inadequate or Imprecise Result or Readings	05/08/2022	05/08/2022
							F1903	Device Explantation	A0908	Incorrect, Inadequate or Imprecise Result or Readings	05/08/2022	05/08/2022
					CT233	Surgical	E2403	No Clinical Signs, Symptoms or Conditions	A0908	Incorrect, Inadequate or Imprecise Result or Readings	05/08/2022	05/08/2022
							F08	Hospitalization or Prolonged Hospitalization	A0908	Incorrect, Inadequate or Imprecise Result or Readings	05/08/2022	05/08/2022
							F1903	Device Explantation	A0908	Incorrect, Inadequate or Imprecise Result or Readings	05/08/2022	05/08/2022
					CT2406	Implantable	E2403	No Clinical Signs, Symptoms or Conditions	A0908	Incorrect, Inadequate or Imprecise Result or Readings	05/08/2022	05/08/2022
							F08	Hospitalization or Prolonged Hospitalization	A0908	Incorrect, Inadequate or Imprecise Result or Readings	05/08/2022	05/08/2022
							F1903	Device Explantation	A0908	Incorrect, Inadequate or Imprecise Result or Readings	05/08/2022	05/08/2022
					CT2407	Active implantable	E2403	No Clinical Signs, Symptoms or Conditions	A0908	Incorrect, Inadequate or Imprecise Result or Readings	05/08/2022	05/08/2022

							F08	Hospitalization or Prolonged Hospitalization	A0908	Incorrect, Inadequate or Imprecise Result or Readings	05/08/2022	05/08/2022
							F1903	Device Explantation	A0908	Incorrect, Inadequate or Imprecise Result or Readings	05/08/2022	05/08/2022
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A0908	Incorrect, Inadequate or Imprecise Result or Readings	05/08/2022	05/08/2022
							F08	Hospitalization or Prolonged Hospitalization	A0908	Incorrect, Inadequate or Imprecise Result or Readings	05/08/2022	05/08/2022
							F1903	Device Explantation	A0908	Incorrect, Inadequate or Imprecise Result or Readings	05/08/2022	05/08/2022
					CT2727	Active implant devices/ancillaries	E2403	No Clinical Signs, Symptoms or Conditions	A0908	Incorrect, Inadequate or Imprecise Result or Readings	05/08/2022	05/08/2022
							F08	Hospitalization or Prolonged Hospitalization	A0908	Incorrect, Inadequate or Imprecise Result or Readings	05/08/2022	05/08/2022
							F1903	Device Explantation	A0908	Incorrect, Inadequate or Imprecise Result or Readings	05/08/2022	05/08/2022
					CT2750	Implantable cardiac electrical stimulators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A0908	Incorrect, Inadequate or Imprecise Result or Readings	05/08/2022	05/08/2022
							F08	Hospitalization or Prolonged Hospitalization	A0908	Incorrect, Inadequate or Imprecise Result or Readings	05/08/2022	05/08/2022
							F1903	Device Explantation	A0908	Incorrect, Inadequate or Imprecise Result or Readings	05/08/2022	05/08/2022
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A0908	Incorrect, Inadequate or Imprecise Result or Readings	05/08/2022	05/08/2022
							F08	Hospitalization or Prolonged Hospitalization	A0908	Incorrect, Inadequate or Imprecise Result or Readings	05/08/2022	05/08/2022
							F1903	Device Explantation	A0908	Incorrect, Inadequate or Imprecise Result or Readings	05/08/2022	05/08/2022
					CT2985	Active	E2403	No Clinical Signs, Symptoms or Conditions	A0908	Incorrect, Inadequate or Imprecise Result or Readings	05/08/2022	05/08/2022
							F08	Hospitalization or Prolonged Hospitalization	A0908	Incorrect, Inadequate or Imprecise Result or Readings	05/08/2022	05/08/2022
							F1903	Device Explantation	A0908	Incorrect, Inadequate or Imprecise Result or Readings	05/08/2022	05/08/2022
					CT3001	Implantable metal	E2403	No Clinical Signs, Symptoms or Conditions	A0908	Incorrect, Inadequate or Imprecise Result or Readings	05/08/2022	05/08/2022
							F08	Hospitalization or Prolonged Hospitalization	A0908	Incorrect, Inadequate or Imprecise Result or Readings	05/08/2022	05/08/2022
							F1903	Device Explantation	A0908	Incorrect, Inadequate or Imprecise Result or Readings	05/08/2022	05/08/2022
					CT3002	Implantable synthetic polymer	E2403	No Clinical Signs, Symptoms or Conditions	A0908	Incorrect, Inadequate or Imprecise Result or Readings	05/08/2022	05/08/2022
							F08	Hospitalization or Prolonged Hospitalization	A0908	Incorrect, Inadequate or Imprecise Result or Readings	05/08/2022	05/08/2022
							F1903	Device Explantation	A0908	Incorrect, Inadequate or Imprecise Result or Readings	05/08/2022	05/08/2022
					CT3006	Implantable inorganic material	E2403	No Clinical Signs, Symptoms or Conditions	A0908	Incorrect, Inadequate or Imprecise Result or Readings	05/08/2022	05/08/2022
							F08	Hospitalization or Prolonged Hospitalization	A0908	Incorrect, Inadequate or Imprecise Result or Readings	05/08/2022	05/08/2022
							F1903	Device Explantation	A0908	Incorrect, Inadequate or Imprecise Result or Readings	05/08/2022	05/08/2022
					CT314	Electrically powered	E2403	No Clinical Signs, Symptoms or Conditions	A0908	Incorrect, Inadequate or Imprecise Result or Readings	05/08/2022	05/08/2022
							F08	Hospitalization or Prolonged Hospitalization	A0908	Incorrect, Inadequate or Imprecise Result or Readings	05/08/2022	05/08/2022
							F1903	Device Explantation	A0908	Incorrect, Inadequate or Imprecise Result or Readings	05/08/2022	05/08/2022
					CT321	Long-term surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A0908	Incorrect, Inadequate or Imprecise Result or Readings	05/08/2022	05/08/2022
							F08	Hospitalization or Prolonged Hospitalization	A0908	Incorrect, Inadequate or Imprecise Result or Readings	05/08/2022	05/08/2022
							F1903	Device Explantation	A0908	Incorrect, Inadequate or Imprecise Result or Readings	05/08/2022	05/08/2022
					CT449	Generators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A0908	Incorrect, Inadequate or Imprecise Result or Readings	05/08/2022	05/08/2022
							F08	Hospitalization or Prolonged Hospitalization	A0908	Incorrect, Inadequate or Imprecise Result or Readings	05/08/2022	05/08/2022
							F1903	Device Explantation	A0908	Incorrect, Inadequate or Imprecise Result or Readings	05/08/2022	05/08/2022
					CT981	Single use	E2403	No Clinical Signs, Symptoms or Conditions	A0908	Incorrect, Inadequate or Imprecise Result or Readings	05/08/2022	05/08/2022
							F08	Hospitalization or Prolonged Hospitalization	A0908	Incorrect, Inadequate or Imprecise Result or Readings	05/08/2022	05/08/2022
							F1903	Device Explantation	A0908	Incorrect, Inadequate or Imprecise Result or Readings	05/08/2022	05/08/2022
					CT983	Surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A0908	Incorrect, Inadequate or Imprecise Result or Readings	05/08/2022	05/08/2022
							F08	Hospitalization or Prolonged Hospitalization	A0908	Incorrect, Inadequate or Imprecise Result or Readings	05/08/2022	05/08/2022
							F1903	Device Explantation	A0908	Incorrect, Inadequate or Imprecise Result or Readings	05/08/2022	05/08/2022
7485	2022	All other reportable incidents	63188	Left-ventricular wireless pacing system sensor/transmitter	CT141	Electrodes/Leads and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A0701	Capturing Problem		
					CT145	Cardiovascular devices	E2403	No Clinical Signs, Symptoms or Conditions	A0701	Capturing Problem		
					CT1577	Wireless monitoring system transmitters	E2403	No Clinical Signs, Symptoms or Conditions	A0701	Capturing Problem		
					CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A0701	Capturing Problem		
					CT2317	Transmitters/Receivers/Transceivers	E2403	No Clinical Signs, Symptoms or Conditions	A0701	Capturing Problem		
					CT233	Surgical	E2403	No Clinical Signs, Symptoms or Conditions	A0701	Capturing Problem		
					CT2406	Implantable	E2403	No Clinical Signs, Symptoms or Conditions	A0701	Capturing Problem		
					CT2407	Active implantable	E2403	No Clinical Signs, Symptoms or Conditions	A0701	Capturing Problem		
					CT255	Transducers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A0701	Capturing Problem		
					CT2726	Physiological monitoring/measuring devices	E2403	No Clinical Signs, Symptoms or Conditions	A0701	Capturing Problem		
					CT2727	Active implant devices/ancillaries	E2403	No Clinical Signs, Symptoms or Conditions	A0701	Capturing Problem		
					CT2750	Implantable cardiac electrical stimulators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A0701	Capturing Problem		
					CT2885	Physiological monitoring/measuring system transmitters/receivers/transceivers	E2403	No Clinical Signs, Symptoms or Conditions	A0701	Capturing Problem		
					CT2985	Active	E2403	No Clinical Signs, Symptoms or Conditions	A0701	Capturing Problem		

					CT3001	Implantable metal	E2403	No Clinical Signs, Symptoms or Conditions	A0701	Capturing Problem		
					CT3002	Implantable synthetic polymer	E2403	No Clinical Signs, Symptoms or Conditions	A0701	Capturing Problem		
					CT3006	Implantable inorganic material	E2403	No Clinical Signs, Symptoms or Conditions	A0701	Capturing Problem		
					CT314	Electrically powered	E2403	No Clinical Signs, Symptoms or Conditions	A0701	Capturing Problem		
					CT321	Long-term surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A0701	Capturing Problem		
					CT384	Sensors and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A0701	Capturing Problem		
					CT449	Generators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A0701	Capturing Problem		
					CT981	Single use	E2403	No Clinical Signs, Symptoms or Conditions	A0701	Capturing Problem		
					CT983	Surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A0701	Capturing Problem		
					CT994	Sensors	E2403	No Clinical Signs, Symptoms or Conditions	A0701	Capturing Problem		
7486	2022	All other reportable incidents	35223	Endocardial/interventricular septal pacing lead	CT1684	Cardiac pacemakers and associated devices	E0112	Dizziness	A0401	Break	04/01/2022	04/01/2022
									A040101	Fracture	04/01/2022	04/01/2022
									A0712	Pacing Problem	04/01/2022	04/01/2022
									A090801	Signal Artifact/Noise	04/01/2022	04/01/2022
							E2403	No Clinical Signs, Symptoms or Conditions	A0401	Break	04/01/2022	04/01/2022
									A040101	Fracture	04/01/2022	04/01/2022
									A0712	Pacing Problem	04/01/2022	04/01/2022
									A090801	Signal Artifact/Noise	04/01/2022	04/01/2022
							F11	Non-Serious Injury/ Illness/ Impairment	A0401	Break	04/01/2022	04/01/2022
									A040101	Fracture	04/01/2022	04/01/2022
									A0712	Pacing Problem	04/01/2022	04/01/2022
									A090801	Signal Artifact/Noise	04/01/2022	04/01/2022
							F1905	Device Revision or Replacement	A0401	Break	04/01/2022	04/01/2022
									A040101	Fracture	04/01/2022	04/01/2022
									A0712	Pacing Problem	04/01/2022	04/01/2022
									A090801	Signal Artifact/Noise	04/01/2022	04/01/2022
7487	2022	All other reportable incidents	47265	Dual-chamber implantable pacemaker, rate-responsive	CT141	Electrodes/Leads and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070909	Oversensing		
							F26	No Health Consequences or Impact	A070909	Oversensing		
					CT145	Cardiovascular devices	E2403	No Clinical Signs, Symptoms or Conditions	A070909	Oversensing		
							F26	No Health Consequences or Impact	A070909	Oversensing		
					CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070909	Oversensing		
							F26	No Health Consequences or Impact	A070909	Oversensing		
					CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A070909	Oversensing		
							F26	No Health Consequences or Impact	A070909	Oversensing		
					CT233	Surgical	E2403	No Clinical Signs, Symptoms or Conditions	A070909	Oversensing		
							F26	No Health Consequences or Impact	A070909	Oversensing		
					CT2406	Implantable	E2403	No Clinical Signs, Symptoms or Conditions	A070909	Oversensing		
							F26	No Health Consequences or Impact	A070909	Oversensing		
					CT2407	Active implantable	E2403	No Clinical Signs, Symptoms or Conditions	A070909	Oversensing		
							F26	No Health Consequences or Impact	A070909	Oversensing		
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A070909	Oversensing		
							F26	No Health Consequences or Impact	A070909	Oversensing		
					CT2727	Active implant devices/ancillaries	E2403	No Clinical Signs, Symptoms or Conditions	A070909	Oversensing		
							F26	No Health Consequences or Impact	A070909	Oversensing		
					CT2750	Implantable cardiac electrical stimulators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070909	Oversensing		
							F26	No Health Consequences or Impact	A070909	Oversensing		
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A070909	Oversensing		
							F26	No Health Consequences or Impact	A070909	Oversensing		
					CT2985	Active	E2403	No Clinical Signs, Symptoms or Conditions	A070909	Oversensing		
							F26	No Health Consequences or Impact	A070909	Oversensing		
					CT3001	Implantable metal	E2403	No Clinical Signs, Symptoms or Conditions	A070909	Oversensing		
							F26	No Health Consequences or Impact	A070909	Oversensing		
					CT3002	Implantable synthetic polymer	E2403	No Clinical Signs, Symptoms or Conditions	A070909	Oversensing		
							F26	No Health Consequences or Impact	A070909	Oversensing		
					CT3006	Implantable inorganic material	E2403	No Clinical Signs, Symptoms or Conditions	A070909	Oversensing		
							F26	No Health Consequences or Impact	A070909	Oversensing		
					CT314	Electrically powered	E2403	No Clinical Signs, Symptoms or Conditions	A070909	Oversensing		
							F26	No Health Consequences or Impact	A070909	Oversensing		
					CT321	Long-term surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A070909	Oversensing		
							F26	No Health Consequences or Impact	A070909	Oversensing		
					CT449	Generators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070909	Oversensing		
							F26	No Health Consequences or Impact	A070909	Oversensing		
					CT981	Single use	E2403	No Clinical Signs, Symptoms or Conditions	A070909	Oversensing		
							F26	No Health Consequences or Impact	A070909	Oversensing		
					CT983	Surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A070909	Oversensing		
							F26	No Health Consequences or Impact	A070909	Oversensing		
7488	2022	All other reportable incidents	35039	Epicardial pacing lead	CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070901	Decreased Sensitivity	10/08/2022	10/08/2022
									A070909	Oversensing	10/08/2022	10/08/2022
									A090801	Signal Artifact/Noise	10/08/2022	10/08/2022
							F08	Hospitalization or Prolonged Hospitalization	A070901	Decreased Sensitivity	10/08/2022	10/08/2022
									A070909	Oversensing	10/08/2022	10/08/2022
									A090801	Signal Artifact/Noise	10/08/2022	10/08/2022
							F12	Serious Injury/ Illness/ Impairment	A070901	Decreased Sensitivity	10/08/2022	10/08/2022
									A070909	Oversensing	10/08/2022	10/08/2022
									A090801	Signal Artifact/Noise	10/08/2022	10/08/2022
							F1905	Device Revision or Replacement	A070901	Decreased Sensitivity	10/08/2022	10/08/2022
									A070909	Oversensing	10/08/2022	10/08/2022
									A090801	Signal Artifact/Noise	10/08/2022	10/08/2022
							F23	Unexpected Medical Intervention	A070901	Decreased Sensitivity	10/08/2022	10/08/2022
									A070909	Oversensing	10/08/2022	10/08/2022
									A090801	Signal Artifact/Noise	10/08/2022	10/08/2022
7489	2022	All other reportable incidents	35852	Single-chamber implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E0601	Arrhythmia	A070909	Oversensing		
									A070912	Incorrect Interpretation of Signal		
									A071302	Inappropriate/Inadequate Shock/Stimulation		
							E2403	No Clinical Signs, Symptoms or Conditions	A070909	Oversensing		
									A070912	Incorrect Interpretation of Signal		
									A071302	Inappropriate/Inadequate Shock/Stimulation		
							F26	No Health Consequences or Impact	A070909	Oversensing		

									A070912	Incorrect Interpretation of Signal		
									A071302	Inappropriate/Inadequate Shock/Stimulation		
7490	2022	All other reportable incidents	37265	Dual-chamber implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
7491	2022	All other reportable incidents	47265	Dual-chamber implantable pacemaker, rate-responsive	CT141	Electrodes/Leads and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem	01/11/2022	01/11/2022
									A070504	Premature Discharge of Battery	01/11/2022	01/11/2022
							F08	Hospitalization or Prolonged Hospitalization	A0705	Battery Problem	01/11/2022	01/11/2022
									A070504	Premature Discharge of Battery	01/11/2022	01/11/2022
							F1905	Device Revision or Replacement	A0705	Battery Problem	01/11/2022	01/11/2022
									A070504	Premature Discharge of Battery	01/11/2022	01/11/2022
					CT145	Cardiovascular devices	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem	01/11/2022	01/11/2022
									A070504	Premature Discharge of Battery	01/11/2022	01/11/2022
							F08	Hospitalization or Prolonged Hospitalization	A0705	Battery Problem	01/11/2022	01/11/2022
									A070504	Premature Discharge of Battery	01/11/2022	01/11/2022
							F1905	Device Revision or Replacement	A0705	Battery Problem	01/11/2022	01/11/2022
									A070504	Premature Discharge of Battery	01/11/2022	01/11/2022
					CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem	01/11/2022	01/11/2022
									A070504	Premature Discharge of Battery	01/11/2022	01/11/2022
							F08	Hospitalization or Prolonged Hospitalization	A0705	Battery Problem	01/11/2022	01/11/2022
									A070504	Premature Discharge of Battery	01/11/2022	01/11/2022
							F1905	Device Revision or Replacement	A0705	Battery Problem	01/11/2022	01/11/2022
									A070504	Premature Discharge of Battery	01/11/2022	01/11/2022
					CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem	01/11/2022	01/11/2022
									A070504	Premature Discharge of Battery	01/11/2022	01/11/2022
							F08	Hospitalization or Prolonged Hospitalization	A0705	Battery Problem	01/11/2022	01/11/2022
									A070504	Premature Discharge of Battery	01/11/2022	01/11/2022
							F1905	Device Revision or Replacement	A0705	Battery Problem	01/11/2022	01/11/2022
									A070504	Premature Discharge of Battery	01/11/2022	01/11/2022
					CT233	Surgical	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem	01/11/2022	01/11/2022
									A070504	Premature Discharge of Battery	01/11/2022	01/11/2022
							F08	Hospitalization or Prolonged Hospitalization	A0705	Battery Problem	01/11/2022	01/11/2022
									A070504	Premature Discharge of Battery	01/11/2022	01/11/2022
							F1905	Device Revision or Replacement	A0705	Battery Problem	01/11/2022	01/11/2022
									A070504	Premature Discharge of Battery	01/11/2022	01/11/2022
					CT2406	Implantable	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem	01/11/2022	01/11/2022
									A070504	Premature Discharge of Battery	01/11/2022	01/11/2022
							F08	Hospitalization or Prolonged Hospitalization	A0705	Battery Problem	01/11/2022	01/11/2022
									A070504	Premature Discharge of Battery	01/11/2022	01/11/2022
							F1905	Device Revision or Replacement	A0705	Battery Problem	01/11/2022	01/11/2022
									A070504	Premature Discharge of Battery	01/11/2022	01/11/2022
					CT2407	Active implantable	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem	01/11/2022	01/11/2022
									A070504	Premature Discharge of Battery	01/11/2022	01/11/2022
							F08	Hospitalization or Prolonged Hospitalization	A0705	Battery Problem	01/11/2022	01/11/2022
									A070504	Premature Discharge of Battery	01/11/2022	01/11/2022
							F1905	Device Revision or Replacement	A0705	Battery Problem	01/11/2022	01/11/2022
									A070504	Premature Discharge of Battery	01/11/2022	01/11/2022
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem	01/11/2022	01/11/2022
									A070504	Premature Discharge of Battery	01/11/2022	01/11/2022
							F08	Hospitalization or Prolonged Hospitalization	A0705	Battery Problem	01/11/2022	01/11/2022
									A070504	Premature Discharge of Battery	01/11/2022	01/11/2022
							F1905	Device Revision or Replacement	A0705	Battery Problem	01/11/2022	01/11/2022
									A070504	Premature Discharge of Battery	01/11/2022	01/11/2022
					CT2727	Active implant devices/ancillaries	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem	01/11/2022	01/11/2022
									A070504	Premature Discharge of Battery	01/11/2022	01/11/2022
							F08	Hospitalization or Prolonged Hospitalization	A0705	Battery Problem	01/11/2022	01/11/2022
									A070504	Premature Discharge of Battery	01/11/2022	01/11/2022
							F1905	Device Revision or Replacement	A0705	Battery Problem	01/11/2022	01/11/2022
									A070504	Premature Discharge of Battery	01/11/2022	01/11/2022
					CT2750	Implantable cardiac electrical stimulators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem	01/11/2022	01/11/2022
									A070504	Premature Discharge of Battery	01/11/2022	01/11/2022
							F08	Hospitalization or Prolonged Hospitalization	A0705	Battery Problem	01/11/2022	01/11/2022
									A070504	Premature Discharge of Battery	01/11/2022	01/11/2022
							F1905	Device Revision or Replacement	A0705	Battery Problem	01/11/2022	01/11/2022
									A070504	Premature Discharge of Battery	01/11/2022	01/11/2022
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem	01/11/2022	01/11/2022
									A070504	Premature Discharge of Battery	01/11/2022	01/11/2022
							F08	Hospitalization or Prolonged Hospitalization	A0705	Battery Problem	01/11/2022	01/11/2022
									A070504	Premature Discharge of Battery	01/11/2022	01/11/2022
							F1905	Device Revision or Replacement	A0705	Battery Problem	01/11/2022	01/11/2022
									A070504	Premature Discharge of Battery	01/11/2022	01/11/2022
					CT2985	Active	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem	01/11/2022	01/11/2022
									A070504	Premature Discharge of Battery	01/11/2022	01/11/2022
							F08	Hospitalization or Prolonged Hospitalization	A0705	Battery Problem	01/11/2022	01/11/2022
									A070504	Premature Discharge of Battery	01/11/2022	01/11/2022
							F1905	Device Revision or Replacement	A0705	Battery Problem	01/11/2022	01/11/2022
									A070504	Premature Discharge of Battery	01/11/2022	01/11/2022
					CT3001	Implantable metal	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem	01/11/2022	01/11/2022
									A070504	Premature Discharge of Battery	01/11/2022	01/11/2022
							F08	Hospitalization or Prolonged Hospitalization	A0705	Battery Problem	01/11/2022	01/11/2022
									A070504	Premature Discharge of Battery	01/11/2022	01/11/2022
							F1905	Device Revision or Replacement	A0705	Battery Problem	01/11/2022	01/11/2022
									A070504	Premature Discharge of Battery	01/11/2022	01/11/2022
					CT3002	Implantable synthetic polymer	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem	01/11/2022	01/11/2022
									A070504	Premature Discharge of Battery	01/11/2022	01/11/2022
							F08	Hospitalization or Prolonged Hospitalization	A0705	Battery Problem	01/11/2022	01/11/2022
									A070504	Premature Discharge of Battery	01/11/2022	01/11/2022
							F1905	Device Revision or Replacement	A0705	Battery Problem	01/11/2022	01/11/2022
									A070504	Premature Discharge of Battery	01/11/2022	01/11/2022

					CT3006	Implantable inorganic material	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem	01/11/2022	01/11/2022
									A070504	Premature Discharge of Battery	01/11/2022	01/11/2022
							F08	Hospitalization or Prolonged Hospitalization	A0705	Battery Problem	01/11/2022	01/11/2022
									A070504	Premature Discharge of Battery	01/11/2022	01/11/2022
							F1905	Device Revision or Replacement	A0705	Battery Problem	01/11/2022	01/11/2022
									A070504	Premature Discharge of Battery	01/11/2022	01/11/2022
					CT314	Electrically powered	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem	01/11/2022	01/11/2022
									A070504	Premature Discharge of Battery	01/11/2022	01/11/2022
							F08	Hospitalization or Prolonged Hospitalization	A0705	Battery Problem	01/11/2022	01/11/2022
									A070504	Premature Discharge of Battery	01/11/2022	01/11/2022
							F1905	Device Revision or Replacement	A0705	Battery Problem	01/11/2022	01/11/2022
									A070504	Premature Discharge of Battery	01/11/2022	01/11/2022
					CT321	Long-term surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem	01/11/2022	01/11/2022
									A070504	Premature Discharge of Battery	01/11/2022	01/11/2022
							F08	Hospitalization or Prolonged Hospitalization	A0705	Battery Problem	01/11/2022	01/11/2022
									A070504	Premature Discharge of Battery	01/11/2022	01/11/2022
							F1905	Device Revision or Replacement	A0705	Battery Problem	01/11/2022	01/11/2022
									A070504	Premature Discharge of Battery	01/11/2022	01/11/2022
					CT449	Generators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem	01/11/2022	01/11/2022
									A070504	Premature Discharge of Battery	01/11/2022	01/11/2022
							F08	Hospitalization or Prolonged Hospitalization	A0705	Battery Problem	01/11/2022	01/11/2022
									A070504	Premature Discharge of Battery	01/11/2022	01/11/2022
							F1905	Device Revision or Replacement	A0705	Battery Problem	01/11/2022	01/11/2022
									A070504	Premature Discharge of Battery	01/11/2022	01/11/2022
					CT981	Single use	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem	01/11/2022	01/11/2022
									A070504	Premature Discharge of Battery	01/11/2022	01/11/2022
							F08	Hospitalization or Prolonged Hospitalization	A0705	Battery Problem	01/11/2022	01/11/2022
									A070504	Premature Discharge of Battery	01/11/2022	01/11/2022
							F1905	Device Revision or Replacement	A0705	Battery Problem	01/11/2022	01/11/2022
									A070504	Premature Discharge of Battery	01/11/2022	01/11/2022
					CT983	Surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem	01/11/2022	01/11/2022
									A070504	Premature Discharge of Battery	01/11/2022	01/11/2022
							F08	Hospitalization or Prolonged Hospitalization	A0705	Battery Problem	01/11/2022	01/11/2022
									A070504	Premature Discharge of Battery	01/11/2022	01/11/2022
							F1905	Device Revision or Replacement	A0705	Battery Problem	01/11/2022	01/11/2022
									A070504	Premature Discharge of Battery	01/11/2022	01/11/2022
7492	2022	All other reportable incidents	35223	Endocardial/interventricular septal pacing lead	CT1684	Cardiac pacemakers and associated devices	E0112	Dizziness	A070909	Oversensing		
							E2403	No Clinical Signs, Symptoms or Conditions	A070909	Oversensing		
							F1905	Device Revision or Replacement	A070909	Oversensing		
7493	2022	All other reportable incidents	47265	Dual-chamber implantable pacemaker, rate-responsive	CT141	Electrodes/Leads and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
					CT145	Cardiovascular devices	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
					CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
					CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
					CT233	Surgical	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
					CT2406	Implantable	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
					CT2407	Active implantable	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
					CT2727	Active implant devices/ancillaries	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
					CT2750	Implantable cardiac electrical stimulators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
					CT2985	Active	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
					CT3001	Implantable metal	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem		

									A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
					CT3002	Implantable synthetic polymer	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
					CT3006	Implantable inorganic material	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
					CT314	Electrically powered	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
					CT321	Long-term surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
					CT449	Generators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
					CT981	Single use	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
					CT983	Surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
7494	2022	All other reportable incidents	35223	Endocardial/interventricular septal pacing lead	CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A0401	Break		
									A070909	Oversensing		
							F26	No Health Consequences or Impact	A0401	Break		
									A070909	Oversensing		
7495	2022	All other reportable incidents	37265	Dual-chamber implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem	11/01/2022	11/01/2022
									A070504	Premature Discharge of Battery	11/01/2022	11/01/2022
							F08	Hospitalization or Prolonged Hospitalization	A0705	Battery Problem	11/01/2022	11/01/2022
									A070504	Premature Discharge of Battery	11/01/2022	11/01/2022
							F1905	Device Revision or Replacement	A0705	Battery Problem	11/01/2022	11/01/2022
									A070504	Premature Discharge of Battery	11/01/2022	11/01/2022
7496	2022	All other reportable incidents	35852	Single-chamber implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem	18/01/2022	18/01/2022
									A070504	Premature Discharge of Battery	18/01/2022	18/01/2022
							F08	Hospitalization or Prolonged Hospitalization	A0705	Battery Problem	18/01/2022	18/01/2022
									A070504	Premature Discharge of Battery	18/01/2022	18/01/2022
							F1905	Device Revision or Replacement	A0705	Battery Problem	18/01/2022	18/01/2022
									A070504	Premature Discharge of Battery	18/01/2022	18/01/2022
7497	2022	All other reportable incidents	35223	Endocardial/interventricular septal pacing lead	CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A072202	Low Impedance	25/08/2021	
							F26	No Health Consequences or Impact	A072202	Low Impedance	25/08/2021	
7498	2022	All other reportable incidents	47265	Dual-chamber implantable pacemaker, rate-responsive	CT141	Electrodes/Leads and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A0711	Interrogation Problem		
							F26	No Health Consequences or Impact	A0711	Interrogation Problem		
					CT145	Cardiovascular devices	E2403	No Clinical Signs, Symptoms or Conditions	A0711	Interrogation Problem		
							F26	No Health Consequences or Impact	A0711	Interrogation Problem		
					CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A0711	Interrogation Problem		
							F26	No Health Consequences or Impact	A0711	Interrogation Problem		
					CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A0711	Interrogation Problem		
							F26	No Health Consequences or Impact	A0711	Interrogation Problem		
					CT233	Surgical	E2403	No Clinical Signs, Symptoms or Conditions	A0711	Interrogation Problem		
							F26	No Health Consequences or Impact	A0711	Interrogation Problem		
					CT2406	Implantable	E2403	No Clinical Signs, Symptoms or Conditions	A0711	Interrogation Problem		
							F26	No Health Consequences or Impact	A0711	Interrogation Problem		
					CT2407	Active implantable	E2403	No Clinical Signs, Symptoms or Conditions	A0711	Interrogation Problem		
							F26	No Health Consequences or Impact	A0711	Interrogation Problem		
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A0711	Interrogation Problem		
							F26	No Health Consequences or Impact	A0711	Interrogation Problem		
					CT2727	Active implant devices/ancillaries	E2403	No Clinical Signs, Symptoms or Conditions	A0711	Interrogation Problem		
							F26	No Health Consequences or Impact	A0711	Interrogation Problem		
					CT2750	Implantable cardiac electrical stimulators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A0711	Interrogation Problem		
							F26	No Health Consequences or Impact	A0711	Interrogation Problem		
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A0711	Interrogation Problem		
							F26	No Health Consequences or Impact	A0711	Interrogation Problem		
					CT2985	Active	E2403	No Clinical Signs, Symptoms or Conditions	A0711	Interrogation Problem		
							F26	No Health Consequences or Impact	A0711	Interrogation Problem		
					CT3001	Implantable metal	E2403	No Clinical Signs, Symptoms or Conditions	A0711	Interrogation Problem		
							F26	No Health Consequences or Impact	A0711	Interrogation Problem		
					CT3002	Implantable synthetic polymer	E2403	No Clinical Signs, Symptoms or Conditions	A0711	Interrogation Problem		
							F26	No Health Consequences or Impact	A0711	Interrogation Problem		
					CT3006	Implantable inorganic material	E2403	No Clinical Signs, Symptoms or Conditions	A0711	Interrogation Problem		
							F26	No Health Consequences or Impact	A0711	Interrogation Problem		
					CT314	Electrically powered	E2403	No Clinical Signs, Symptoms or Conditions	A0711	Interrogation Problem		
							F26	No Health Consequences or Impact	A0711	Interrogation Problem		
					CT321	Long-term surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A0711	Interrogation Problem		
							F26	No Health Consequences or Impact	A0711	Interrogation Problem		
					CT449	Generators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A0711	Interrogation Problem		
							F26	No Health Consequences or Impact	A0711	Interrogation Problem		
					CT981	Single use	E2403	No Clinical Signs, Symptoms or Conditions	A0711	Interrogation Problem		
							F26	No Health Consequences or Impact	A0711	Interrogation Problem		
					CT983	Surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A0711	Interrogation Problem		

							F26	No Health Consequences or Impact	A0711	Interrogation Problem		
7499	2022	All other reportable incidents	37265	Dual-chamber implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E060110	Ventricular Fibrillation	A071201	Failure to Convert Rhythm	08/12/2021	08/12/2021
									A072201	High Impedance	08/12/2021	08/12/2021
									A0908	Incorrect, Inadequate or Imprecise Result or Readings	08/12/2021	08/12/2021
							E2104	Electric Shock	A071201	Failure to Convert Rhythm	08/12/2021	08/12/2021
									A072201	High Impedance	08/12/2021	08/12/2021
									A0908	Incorrect, Inadequate or Imprecise Result or Readings	08/12/2021	08/12/2021
							E2403	No Clinical Signs, Symptoms or Conditions	A071201	Failure to Convert Rhythm	08/12/2021	08/12/2021
									A072201	High Impedance	08/12/2021	08/12/2021
									A0908	Incorrect, Inadequate or Imprecise Result or Readings	08/12/2021	08/12/2021
							F08	Hospitalization or Prolonged Hospitalization	A071201	Failure to Convert Rhythm	08/12/2021	08/12/2021
									A072201	High Impedance	08/12/2021	08/12/2021
									A0908	Incorrect, Inadequate or Imprecise Result or Readings	08/12/2021	08/12/2021
							F1905	Device Revision or Replacement	A071201	Failure to Convert Rhythm	08/12/2021	08/12/2021
									A072201	High Impedance	08/12/2021	08/12/2021
									A0908	Incorrect, Inadequate or Imprecise Result or Readings	08/12/2021	08/12/2021
							F22	Unexpected Diagnostic Intervention	A071201	Failure to Convert Rhythm	08/12/2021	08/12/2021
									A072201	High Impedance	08/12/2021	08/12/2021
									A0908	Incorrect, Inadequate or Imprecise Result or Readings	08/12/2021	08/12/2021
							F2306	Resuscitation	A071201	Failure to Convert Rhythm	08/12/2021	08/12/2021
									A072201	High Impedance	08/12/2021	08/12/2021
									A0908	Incorrect, Inadequate or Imprecise Result or Readings	08/12/2021	08/12/2021
7500	2022	All other reportable incidents	35852	Single-chamber implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem	19/01/2022	19/01/2022
									A070504	Premature Discharge of Battery	19/01/2022	19/01/2022
							F08	Hospitalization or Prolonged Hospitalization	A0705	Battery Problem	19/01/2022	19/01/2022
									A070504	Premature Discharge of Battery	19/01/2022	19/01/2022
							F1905	Device Revision or Replacement	A0705	Battery Problem	19/01/2022	19/01/2022
									A070504	Premature Discharge of Battery	19/01/2022	19/01/2022
7501	2022	All other reportable incidents	36241	Pacemaker/defibrillator lead	CT1684	Cardiac pacemakers and associated devices	E060110	Ventricular Fibrillation	A071201	Failure to Convert Rhythm	08/12/2021	08/12/2021
									A072201	High Impedance	08/12/2021	08/12/2021
							E2104	Electric Shock	A071201	Failure to Convert Rhythm	08/12/2021	08/12/2021
									A072201	High Impedance	08/12/2021	08/12/2021
							E2403	No Clinical Signs, Symptoms or Conditions	A071201	Failure to Convert Rhythm	08/12/2021	08/12/2021
									A072201	High Impedance	08/12/2021	08/12/2021
							F08	Hospitalization or Prolonged Hospitalization	A071201	Failure to Convert Rhythm	08/12/2021	08/12/2021
									A072201	High Impedance	08/12/2021	08/12/2021
							F1905	Device Revision or Replacement	A071201	Failure to Convert Rhythm	08/12/2021	08/12/2021
									A072201	High Impedance	08/12/2021	08/12/2021
							F22	Unexpected Diagnostic Intervention	A071201	Failure to Convert Rhythm	08/12/2021	08/12/2021
									A072201	High Impedance	08/12/2021	08/12/2021
							F2306	Resuscitation	A071201	Failure to Convert Rhythm	08/12/2021	08/12/2021
									A072201	High Impedance	08/12/2021	08/12/2021
7502	2022	All other reportable incidents	47263	Cardiac resynchronization therapy implantable pacemaker	CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
7503	2022	All other reportable incidents	35852	Single-chamber implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem	14/12/2022	14/12/2022
									A070504	Premature Discharge of Battery	14/12/2022	14/12/2022
							F08	Hospitalization or Prolonged Hospitalization	A0705	Battery Problem	14/12/2022	14/12/2022
									A070504	Premature Discharge of Battery	14/12/2022	14/12/2022
							F1903	Device Explantation	A0705	Battery Problem	14/12/2022	14/12/2022
									A070504	Premature Discharge of Battery	14/12/2022	14/12/2022
7504	2022	All other reportable incidents	35852	Single-chamber implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem	19/12/2022	19/12/2022
									A070504	Premature Discharge of Battery	19/12/2022	19/12/2022
							F08	Hospitalization or Prolonged Hospitalization	A0705	Battery Problem	19/12/2022	19/12/2022
									A070504	Premature Discharge of Battery	19/12/2022	19/12/2022
							F1903	Device Explantation	A0705	Battery Problem	19/12/2022	19/12/2022
									A070504	Premature Discharge of Battery	19/12/2022	19/12/2022
7505	2022	All other reportable incidents	47265	Dual-chamber implantable pacemaker, rate-responsive	CT141	Electrodes/Leads and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070101	Failure to Capture		
							F26	No Health Consequences or Impact	A070101	Failure to Capture		
					CT145	Cardiovascular devices	E2403	No Clinical Signs, Symptoms or Conditions	A070101	Failure to Capture		
							F26	No Health Consequences or Impact	A070101	Failure to Capture		
					CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070101	Failure to Capture		
							F26	No Health Consequences or Impact	A070101	Failure to Capture		
					CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A070101	Failure to Capture		
							F26	No Health Consequences or Impact	A070101	Failure to Capture		
					CT233	Surgical	E2403	No Clinical Signs, Symptoms or Conditions	A070101	Failure to Capture		
							F26	No Health Consequences or Impact	A070101	Failure to Capture		
					CT2406	Implantable	E2403	No Clinical Signs, Symptoms or Conditions	A070101	Failure to Capture		
							F26	No Health Consequences or Impact	A070101	Failure to Capture		
					CT2407	Active implantable	E2403	No Clinical Signs, Symptoms or Conditions	A070101	Failure to Capture		
							F26	No Health Consequences or Impact	A070101	Failure to Capture		
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A070101	Failure to Capture		
							F26	No Health Consequences or Impact	A070101	Failure to Capture		
					CT2727	Active implant devices/ancillaries	E2403	No Clinical Signs, Symptoms or Conditions	A070101	Failure to Capture		
							F26	No Health Consequences or Impact	A070101	Failure to Capture		
					CT2750	Implantable cardiac electrical stimulators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070101	Failure to Capture		
							F26	No Health Consequences or Impact	A070101	Failure to Capture		

					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A070101	Failure to Capture		
							F26	No Health Consequences or Impact	A070101	Failure to Capture		
					CT2985	Active	E2403	No Clinical Signs, Symptoms or Conditions	A070101	Failure to Capture		
							F26	No Health Consequences or Impact	A070101	Failure to Capture		
					CT3001	Implantable metal	E2403	No Clinical Signs, Symptoms or Conditions	A070101	Failure to Capture		
							F26	No Health Consequences or Impact	A070101	Failure to Capture		
					CT3002	Implantable synthetic polymer	E2403	No Clinical Signs, Symptoms or Conditions	A070101	Failure to Capture		
							F26	No Health Consequences or Impact	A070101	Failure to Capture		
					CT3006	Implantable inorganic material	E2403	No Clinical Signs, Symptoms or Conditions	A070101	Failure to Capture		
							F26	No Health Consequences or Impact	A070101	Failure to Capture		
					CT314	Electrically powered	E2403	No Clinical Signs, Symptoms or Conditions	A070101	Failure to Capture		
							F26	No Health Consequences or Impact	A070101	Failure to Capture		
					CT321	Long-term surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A070101	Failure to Capture		
							F26	No Health Consequences or Impact	A070101	Failure to Capture		
					CT449	Generators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070101	Failure to Capture		
							F26	No Health Consequences or Impact	A070101	Failure to Capture		
					CT981	Single use	E2403	No Clinical Signs, Symptoms or Conditions	A070101	Failure to Capture		
							F26	No Health Consequences or Impact	A070101	Failure to Capture		
					CT983	Surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A070101	Failure to Capture		
							F26	No Health Consequences or Impact	A070101	Failure to Capture		
7506	2022	All other reportable incidents	35223	Endocardial/interventricular septal pacing lead	CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A0401	Break		
									A070909	Oversensing		
							F1908	Prolonged Surgery	A0401	Break		
									A070909	Oversensing		
7507	2022	All other reportable incidents	47265	Dual-chamber implantable pacemaker, rate-responsive	CT141	Electrodes/Leads and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070101	Failure to Capture		
									A070909	Oversensing		
							F26	No Health Consequences or Impact	A070101	Failure to Capture		
									A070909	Oversensing		
					CT145	Cardiovascular devices	E2403	No Clinical Signs, Symptoms or Conditions	A070101	Failure to Capture		
									A070909	Oversensing		
							F26	No Health Consequences or Impact	A070101	Failure to Capture		
									A070909	Oversensing		
					CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070101	Failure to Capture		
									A070909	Oversensing		
							F26	No Health Consequences or Impact	A070101	Failure to Capture		
									A070909	Oversensing		
					CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A070101	Failure to Capture		
									A070909	Oversensing		
							F26	No Health Consequences or Impact	A070101	Failure to Capture		
									A070909	Oversensing		
					CT233	Surgical	E2403	No Clinical Signs, Symptoms or Conditions	A070101	Failure to Capture		
									A070909	Oversensing		
							F26	No Health Consequences or Impact	A070101	Failure to Capture		
									A070909	Oversensing		
					CT2406	Implantable	E2403	No Clinical Signs, Symptoms or Conditions	A070101	Failure to Capture		
									A070909	Oversensing		
							F26	No Health Consequences or Impact	A070101	Failure to Capture		
									A070909	Oversensing		
					CT2407	Active implantable	E2403	No Clinical Signs, Symptoms or Conditions	A070101	Failure to Capture		
									A070909	Oversensing		
							F26	No Health Consequences or Impact	A070101	Failure to Capture		
									A070909	Oversensing		
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A070101	Failure to Capture		
									A070909	Oversensing		
							F26	No Health Consequences or Impact	A070101	Failure to Capture		
									A070909	Oversensing		
					CT2727	Active implant devices/ancillaries	E2403	No Clinical Signs, Symptoms or Conditions	A070101	Failure to Capture		
									A070909	Oversensing		
							F26	No Health Consequences or Impact	A070101	Failure to Capture		
									A070909	Oversensing		
					CT2750	Implantable cardiac electrical stimulators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070101	Failure to Capture		
									A070909	Oversensing		
							F26	No Health Consequences or Impact	A070101	Failure to Capture		
									A070909	Oversensing		
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A070101	Failure to Capture		
									A070909	Oversensing		
							F26	No Health Consequences or Impact	A070101	Failure to Capture		
									A070909	Oversensing		
					CT2985	Active	E2403	No Clinical Signs, Symptoms or Conditions	A070101	Failure to Capture		
									A070909	Oversensing		
							F26	No Health Consequences or Impact	A070101	Failure to Capture		
									A070909	Oversensing		
					CT3001	Implantable metal	E2403	No Clinical Signs, Symptoms or Conditions	A070101	Failure to Capture		
									A070909	Oversensing		
							F26	No Health Consequences or Impact	A070101	Failure to Capture		
									A070909	Oversensing		
					CT3002	Implantable synthetic polymer	E2403	No Clinical Signs, Symptoms or Conditions	A070101	Failure to Capture		
									A070909	Oversensing		
							F26	No Health Consequences or Impact	A070101	Failure to Capture		
									A070909	Oversensing		
					CT3006	Implantable inorganic material	E2403	No Clinical Signs, Symptoms or Conditions	A070101	Failure to Capture		
									A070909	Oversensing		
							F26	No Health Consequences or Impact	A070101	Failure to Capture		
									A070909	Oversensing		
					CT314	Electrically powered	E2403	No Clinical Signs, Symptoms or Conditions	A070101	Failure to Capture		
									A070909	Oversensing		
							F26	No Health Consequences or Impact	A070101	Failure to Capture		
									A070909	Oversensing		

					CT321	Long-term surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A070101	Failure to Capture		
									A070909	Oversensing		
							F26	No Health Consequences or Impact	A070101	Failure to Capture		
									A070909	Oversensing		
					CT449	Generators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070101	Failure to Capture		
									A070909	Oversensing		
							F26	No Health Consequences or Impact	A070101	Failure to Capture		
									A070909	Oversensing		
					CT981	Single use	E2403	No Clinical Signs, Symptoms or Conditions	A070101	Failure to Capture		
									A070909	Oversensing		
							F26	No Health Consequences or Impact	A070101	Failure to Capture		
									A070909	Oversensing		
					CT983	Surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A070101	Failure to Capture		
									A070909	Oversensing		
							F26	No Health Consequences or Impact	A070101	Failure to Capture		
									A070909	Oversensing		
7508	2022	All other reportable incidents	35223	Endocardial/interventricular septal pacing lead	CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070102	High Capture Threshold		
									A0709	Device Sensing Problem		
									A0722	Impedance Problem		
							F26	No Health Consequences or Impact	A070102	High Capture Threshold		
									A0709	Device Sensing Problem		
									A0722	Impedance Problem		
7509	2022	All other reportable incidents	35223	Endocardial/interventricular septal pacing lead	CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070909	Oversensing		
									A072001	Electromagnetic Interference		
									A090801	Signal Artifact/Noise		
							F26	No Health Consequences or Impact	A070909	Oversensing		
									A072001	Electromagnetic Interference		
									A090801	Signal Artifact/Noise		
7510	2022	All other reportable incidents	35223	Endocardial/interventricular septal pacing lead	CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A07	Electrical /Electronic Property Problem		
									A070901	Decreased Sensitivity		
									A070910	Undersensing		
									A0712	Pacing Problem		
							F12	Serious Injury/ Illness/ Impairment	A07	Electrical /Electronic Property Problem		
									A070901	Decreased Sensitivity		
									A070910	Undersensing		
									A0712	Pacing Problem		
							F23	Unexpected Medical Intervention	A07	Electrical /Electronic Property Problem		
									A070901	Decreased Sensitivity		
									A070910	Undersensing		
									A0712	Pacing Problem		
7511	2022	All other reportable incidents	47265	Dual-chamber implantable pacemaker, rate-responsive	CT141	Electrodes/Leads and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070909	Oversensing		
									A0712	Pacing Problem		
									A090801	Signal Artifact/Noise		
							F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A070909	Oversensing		
									A0712	Pacing Problem		
									A090801	Signal Artifact/Noise		
					CT145	Cardiovascular devices	E2403	No Clinical Signs, Symptoms or Conditions	A070909	Oversensing		
									A0712	Pacing Problem		
									A090801	Signal Artifact/Noise		
							F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A070909	Oversensing		
									A0712	Pacing Problem		
									A090801	Signal Artifact/Noise		
					CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070909	Oversensing		
									A0712	Pacing Problem		
									A090801	Signal Artifact/Noise		
							F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A070909	Oversensing		
									A0712	Pacing Problem		
									A090801	Signal Artifact/Noise		
					CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A070909	Oversensing		
									A0712	Pacing Problem		
									A090801	Signal Artifact/Noise		
							F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A070909	Oversensing		
									A0712	Pacing Problem		
									A090801	Signal Artifact/Noise		
					CT233	Surgical	E2403	No Clinical Signs, Symptoms or Conditions	A070909	Oversensing		
									A0712	Pacing Problem		
									A090801	Signal Artifact/Noise		
							F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A070909	Oversensing		
									A0712	Pacing Problem		
									A090801	Signal Artifact/Noise		
					CT2406	Implantable	E2403	No Clinical Signs, Symptoms or Conditions	A070909	Oversensing		
									A0712	Pacing Problem		
									A090801	Signal Artifact/Noise		
							F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A070909	Oversensing		
									A0712	Pacing Problem		
									A090801	Signal Artifact/Noise		
					CT2407	Active implantable	E2403	No Clinical Signs, Symptoms or Conditions	A070909	Oversensing		
									A0712	Pacing Problem		
									A090801	Signal Artifact/Noise		
							F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A070909	Oversensing		
									A0712	Pacing Problem		
									A090801	Signal Artifact/Noise		
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A070909	Oversensing		
									A0712	Pacing Problem		
									A090801	Signal Artifact/Noise		
							F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A070909	Oversensing		
									A0712	Pacing Problem		
									A090801	Signal Artifact/Noise		

					CT2727	Active implant devices/ancillaries	E2403	No Clinical Signs, Symptoms or Conditions	A070909	Oversensing		
									A0712	Pacing Problem		
									A090801	Signal Artifact/Noise		
							F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A070909	Oversensing		
									A0712	Pacing Problem		
									A090801	Signal Artifact/Noise		
					CT2750	Implantable cardiac electrical stimulators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070909	Oversensing		
									A0712	Pacing Problem		
									A090801	Signal Artifact/Noise		
							F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A070909	Oversensing		
									A0712	Pacing Problem		
									A090801	Signal Artifact/Noise		
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A070909	Oversensing		
									A0712	Pacing Problem		
									A090801	Signal Artifact/Noise		
							F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A070909	Oversensing		
									A0712	Pacing Problem		
									A090801	Signal Artifact/Noise		
					CT2985	Active	E2403	No Clinical Signs, Symptoms or Conditions	A070909	Oversensing		
									A0712	Pacing Problem		
									A090801	Signal Artifact/Noise		
							F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A070909	Oversensing		
									A0712	Pacing Problem		
									A090801	Signal Artifact/Noise		
					CT3001	Implantable metal	E2403	No Clinical Signs, Symptoms or Conditions	A070909	Oversensing		
									A0712	Pacing Problem		
									A090801	Signal Artifact/Noise		
							F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A070909	Oversensing		
									A0712	Pacing Problem		
									A090801	Signal Artifact/Noise		
					CT3002	Implantable synthetic polymer	E2403	No Clinical Signs, Symptoms or Conditions	A070909	Oversensing		
									A0712	Pacing Problem		
									A090801	Signal Artifact/Noise		
							F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A070909	Oversensing		
									A0712	Pacing Problem		
									A090801	Signal Artifact/Noise		
					CT3006	Implantable inorganic material	E2403	No Clinical Signs, Symptoms or Conditions	A070909	Oversensing		
									A0712	Pacing Problem		
									A090801	Signal Artifact/Noise		
							F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A070909	Oversensing		
									A0712	Pacing Problem		
									A090801	Signal Artifact/Noise		
					CT314	Electrically powered	E2403	No Clinical Signs, Symptoms or Conditions	A070909	Oversensing		
									A0712	Pacing Problem		
									A090801	Signal Artifact/Noise		
							F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A070909	Oversensing		
									A0712	Pacing Problem		
									A090801	Signal Artifact/Noise		
					CT321	Long-term surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A070909	Oversensing		
									A0712	Pacing Problem		
									A090801	Signal Artifact/Noise		
							F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A070909	Oversensing		
									A0712	Pacing Problem		
									A090801	Signal Artifact/Noise		
					CT449	Generators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070909	Oversensing		
									A0712	Pacing Problem		
									A090801	Signal Artifact/Noise		
							F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A070909	Oversensing		
									A0712	Pacing Problem		
									A090801	Signal Artifact/Noise		
					CT981	Single use	E2403	No Clinical Signs, Symptoms or Conditions	A070909	Oversensing		
									A0712	Pacing Problem		
									A090801	Signal Artifact/Noise		
							F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A070909	Oversensing		
									A0712	Pacing Problem		
									A090801	Signal Artifact/Noise		
					CT983	Surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A070909	Oversensing		
									A0712	Pacing Problem		
									A090801	Signal Artifact/Noise		
							F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A070909	Oversensing		
									A0712	Pacing Problem		
									A090801	Signal Artifact/Noise		
7512	2022	All other reportable incidents	35223	Endocardial/interventricular septal pacing lead	CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070910	Undersensing		
									A072001	Electromagnetic Interference		
									A090801	Signal Artifact/Noise		
							F26	No Health Consequences or Impact	A070910	Undersensing		
									A072001	Electromagnetic Interference		
									A090801	Signal Artifact/Noise		
7513	2022	All other reportable incidents	36241	Pacemaker/defibrillator lead	CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A0722	Impedance Problem		
									A072201	High Impedance		
							F26	No Health Consequences or Impact	A0722	Impedance Problem		
									A072201	High Impedance		
7514	2022	All other reportable incidents	47263	Cardiac resynchronization therapy implantable pacemaker	CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A0908	Incorrect, Inadequate or Imprecise Result or Readings	06/09/2022	06/09/2022
							F08	Hospitalization or Prolonged Hospitalization	A0908	Incorrect, Inadequate or Imprecise Result or Readings	06/09/2022	06/09/2022
							F1905	Device Revision or Replacement	A0908	Incorrect, Inadequate or Imprecise Result or Readings	06/09/2022	06/09/2022

					CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A0908	Incorrect, Inadequate or Imprecise Result or Readings	06/09/2022	06/09/2022
							F08	Hospitalization or Prolonged Hospitalization	A0908	Incorrect, Inadequate or Imprecise Result or Readings	06/09/2022	06/09/2022
							F1905	Device Revision or Replacement	A0908	Incorrect, Inadequate or Imprecise Result or Readings	06/09/2022	06/09/2022
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A0908	Incorrect, Inadequate or Imprecise Result or Readings	06/09/2022	06/09/2022
							F08	Hospitalization or Prolonged Hospitalization	A0908	Incorrect, Inadequate or Imprecise Result or Readings	06/09/2022	06/09/2022
							F1905	Device Revision or Replacement	A0908	Incorrect, Inadequate or Imprecise Result or Readings	06/09/2022	06/09/2022
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A0908	Incorrect, Inadequate or Imprecise Result or Readings	06/09/2022	06/09/2022
							F08	Hospitalization or Prolonged Hospitalization	A0908	Incorrect, Inadequate or Imprecise Result or Readings	06/09/2022	06/09/2022
							F1905	Device Revision or Replacement	A0908	Incorrect, Inadequate or Imprecise Result or Readings	06/09/2022	06/09/2022
7515	2022	All other reportable incidents	35852	Single-chamber implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem	27/03/2022	27/03/2022
									A070504	Premature Discharge of Battery	27/03/2022	27/03/2022
							F08	Hospitalization or Prolonged Hospitalization	A0705	Battery Problem	27/03/2022	27/03/2022
									A070504	Premature Discharge of Battery	27/03/2022	27/03/2022
							F1903	Device Explantation	A0705	Battery Problem	27/03/2022	27/03/2022
									A070504	Premature Discharge of Battery	27/03/2022	27/03/2022
7516	2022	All other reportable incidents	35852	Single-chamber implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem	04/04/2022	04/04/2022
									A070504	Premature Discharge of Battery	04/04/2022	04/04/2022
							F08	Hospitalization or Prolonged Hospitalization	A0705	Battery Problem	04/04/2022	04/04/2022
									A070504	Premature Discharge of Battery	04/04/2022	04/04/2022
							F1905	Device Revision or Replacement	A0705	Battery Problem	04/04/2022	04/04/2022
									A070504	Premature Discharge of Battery	04/04/2022	04/04/2022
7517	2022	All other reportable incidents	35852	Single-chamber implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem	17/01/2022	17/01/2022
									A070504	Premature Discharge of Battery	17/01/2022	17/01/2022
							F08	Hospitalization or Prolonged Hospitalization	A0705	Battery Problem	17/01/2022	17/01/2022
									A070504	Premature Discharge of Battery	17/01/2022	17/01/2022
							F1905	Device Revision or Replacement	A0705	Battery Problem	17/01/2022	17/01/2022
									A070504	Premature Discharge of Battery	17/01/2022	17/01/2022
7518	2022	All other reportable incidents	35852	Single-chamber implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem	20/01/2022	20/01/2022
									A070504	Premature Discharge of Battery	20/01/2022	20/01/2022
							F08	Hospitalization or Prolonged Hospitalization	A0705	Battery Problem	20/01/2022	20/01/2022
									A070504	Premature Discharge of Battery	20/01/2022	20/01/2022
							F1905	Device Revision or Replacement	A0705	Battery Problem	20/01/2022	20/01/2022
									A070504	Premature Discharge of Battery	20/01/2022	20/01/2022
7519	2022	All other reportable incidents	37265	Dual-chamber implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	29/09/2022	29/09/2022
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	29/09/2022	29/09/2022
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	29/09/2022	29/09/2022
7520	2022	All other reportable incidents	47265	Dual-chamber implantable pacemaker, rate-responsive	CT141	Electrodes/Leads and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem	24/12/2021	10/03/2022
									A070504	Premature Discharge of Battery	24/12/2021	10/03/2022
							F08	Hospitalization or Prolonged Hospitalization	A0705	Battery Problem	24/12/2021	10/03/2022
									A070504	Premature Discharge of Battery	24/12/2021	10/03/2022
							F1905	Device Revision or Replacement	A0705	Battery Problem	24/12/2021	10/03/2022
									A070504	Premature Discharge of Battery	24/12/2021	10/03/2022
					CT145	Cardiovascular devices	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem	24/12/2021	10/03/2022
									A070504	Premature Discharge of Battery	24/12/2021	10/03/2022
							F08	Hospitalization or Prolonged Hospitalization	A0705	Battery Problem	24/12/2021	10/03/2022
									A070504	Premature Discharge of Battery	24/12/2021	10/03/2022
							F1905	Device Revision or Replacement	A0705	Battery Problem	24/12/2021	10/03/2022
									A070504	Premature Discharge of Battery	24/12/2021	10/03/2022
					CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem	24/12/2021	10/03/2022
									A070504	Premature Discharge of Battery	24/12/2021	10/03/2022
							F08	Hospitalization or Prolonged Hospitalization	A0705	Battery Problem	24/12/2021	10/03/2022
									A070504	Premature Discharge of Battery	24/12/2021	10/03/2022
							F1905	Device Revision or Replacement	A0705	Battery Problem	24/12/2021	10/03/2022
									A070504	Premature Discharge of Battery	24/12/2021	10/03/2022
					CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem	24/12/2021	10/03/2022
									A070504	Premature Discharge of Battery	24/12/2021	10/03/2022
							F08	Hospitalization or Prolonged Hospitalization	A0705	Battery Problem	24/12/2021	10/03/2022
									A070504	Premature Discharge of Battery	24/12/2021	10/03/2022
							F1905	Device Revision or Replacement	A0705	Battery Problem	24/12/2021	10/03/2022
									A070504	Premature Discharge of Battery	24/12/2021	10/03/2022
					CT233	Surgical	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem	24/12/2021	10/03/2022
									A070504	Premature Discharge of Battery	24/12/2021	10/03/2022
							F08	Hospitalization or Prolonged Hospitalization	A0705	Battery Problem	24/12/2021	10/03/2022
									A070504	Premature Discharge of Battery	24/12/2021	10/03/2022
							F1905	Device Revision or Replacement	A0705	Battery Problem	24/12/2021	10/03/2022
									A070504	Premature Discharge of Battery	24/12/2021	10/03/2022
					CT2406	Implantable	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem	24/12/2021	10/03/2022
									A070504	Premature Discharge of Battery	24/12/2021	10/03/2022
							F08	Hospitalization or Prolonged Hospitalization	A0705	Battery Problem	24/12/2021	10/03/2022
									A070504	Premature Discharge of Battery	24/12/2021	10/03/2022
							F1905	Device Revision or Replacement	A0705	Battery Problem	24/12/2021	10/03/2022
									A070504	Premature Discharge of Battery	24/12/2021	10/03/2022
					CT2407	Active implantable	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem	24/12/2021	10/03/2022
									A070504	Premature Discharge of Battery	24/12/2021	10/03/2022
							F08	Hospitalization or Prolonged Hospitalization	A0705	Battery Problem	24/12/2021	10/03/2022
									A070504	Premature Discharge of Battery	24/12/2021	10/03/2022
							F1905	Device Revision or Replacement	A0705	Battery Problem	24/12/2021	10/03/2022
									A070504	Premature Discharge of Battery	24/12/2021	10/03/2022
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem	24/12/2021	10/03/2022
									A070504	Premature Discharge of Battery	24/12/2021	10/03/2022
							F08	Hospitalization or Prolonged Hospitalization	A0705	Battery Problem	24/12/2021	10/03/2022

									A070504	Premature Discharge of Battery	24/12/2021	10/03/2022
							F1905	Device Revision or Replacement	A0705	Battery Problem	24/12/2021	10/03/2022
									A070504	Premature Discharge of Battery	24/12/2021	10/03/2022
					CT2727	Active implant devices/ancillaries	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem	24/12/2021	10/03/2022
									A070504	Premature Discharge of Battery	24/12/2021	10/03/2022
							F08	Hospitalization or Prolonged Hospitalization	A0705	Battery Problem	24/12/2021	10/03/2022
									A070504	Premature Discharge of Battery	24/12/2021	10/03/2022
							F1905	Device Revision or Replacement	A0705	Battery Problem	24/12/2021	10/03/2022
									A070504	Premature Discharge of Battery	24/12/2021	10/03/2022
					CT2750	Implantable cardiac electrical stimulators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem	24/12/2021	10/03/2022
									A070504	Premature Discharge of Battery	24/12/2021	10/03/2022
							F08	Hospitalization or Prolonged Hospitalization	A0705	Battery Problem	24/12/2021	10/03/2022
									A070504	Premature Discharge of Battery	24/12/2021	10/03/2022
							F1905	Device Revision or Replacement	A0705	Battery Problem	24/12/2021	10/03/2022
									A070504	Premature Discharge of Battery	24/12/2021	10/03/2022
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem	24/12/2021	10/03/2022
									A070504	Premature Discharge of Battery	24/12/2021	10/03/2022
							F08	Hospitalization or Prolonged Hospitalization	A0705	Battery Problem	24/12/2021	10/03/2022
									A070504	Premature Discharge of Battery	24/12/2021	10/03/2022
							F1905	Device Revision or Replacement	A0705	Battery Problem	24/12/2021	10/03/2022
									A070504	Premature Discharge of Battery	24/12/2021	10/03/2022
					CT2985	Active	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem	24/12/2021	10/03/2022
									A070504	Premature Discharge of Battery	24/12/2021	10/03/2022
							F08	Hospitalization or Prolonged Hospitalization	A0705	Battery Problem	24/12/2021	10/03/2022
									A070504	Premature Discharge of Battery	24/12/2021	10/03/2022
							F1905	Device Revision or Replacement	A0705	Battery Problem	24/12/2021	10/03/2022
									A070504	Premature Discharge of Battery	24/12/2021	10/03/2022
					CT3001	Implantable metal	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem	24/12/2021	10/03/2022
									A070504	Premature Discharge of Battery	24/12/2021	10/03/2022
							F08	Hospitalization or Prolonged Hospitalization	A0705	Battery Problem	24/12/2021	10/03/2022
									A070504	Premature Discharge of Battery	24/12/2021	10/03/2022
							F1905	Device Revision or Replacement	A0705	Battery Problem	24/12/2021	10/03/2022
									A070504	Premature Discharge of Battery	24/12/2021	10/03/2022
					CT3002	Implantable synthetic polymer	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem	24/12/2021	10/03/2022
									A070504	Premature Discharge of Battery	24/12/2021	10/03/2022
							F08	Hospitalization or Prolonged Hospitalization	A0705	Battery Problem	24/12/2021	10/03/2022
									A070504	Premature Discharge of Battery	24/12/2021	10/03/2022
							F1905	Device Revision or Replacement	A0705	Battery Problem	24/12/2021	10/03/2022
									A070504	Premature Discharge of Battery	24/12/2021	10/03/2022
					CT3006	Implantable inorganic material	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem	24/12/2021	10/03/2022
									A070504	Premature Discharge of Battery	24/12/2021	10/03/2022
							F08	Hospitalization or Prolonged Hospitalization	A0705	Battery Problem	24/12/2021	10/03/2022
									A070504	Premature Discharge of Battery	24/12/2021	10/03/2022
							F1905	Device Revision or Replacement	A0705	Battery Problem	24/12/2021	10/03/2022
									A070504	Premature Discharge of Battery	24/12/2021	10/03/2022
					CT314	Electrically powered	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem	24/12/2021	10/03/2022
									A070504	Premature Discharge of Battery	24/12/2021	10/03/2022
							F08	Hospitalization or Prolonged Hospitalization	A0705	Battery Problem	24/12/2021	10/03/2022
									A070504	Premature Discharge of Battery	24/12/2021	10/03/2022
							F1905	Device Revision or Replacement	A0705	Battery Problem	24/12/2021	10/03/2022
									A070504	Premature Discharge of Battery	24/12/2021	10/03/2022
					CT321	Long-term surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem	24/12/2021	10/03/2022
									A070504	Premature Discharge of Battery	24/12/2021	10/03/2022
							F08	Hospitalization or Prolonged Hospitalization	A0705	Battery Problem	24/12/2021	10/03/2022
									A070504	Premature Discharge of Battery	24/12/2021	10/03/2022
							F1905	Device Revision or Replacement	A0705	Battery Problem	24/12/2021	10/03/2022
									A070504	Premature Discharge of Battery	24/12/2021	10/03/2022
					CT449	Generators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem	24/12/2021	10/03/2022
									A070504	Premature Discharge of Battery	24/12/2021	10/03/2022
							F08	Hospitalization or Prolonged Hospitalization	A0705	Battery Problem	24/12/2021	10/03/2022
									A070504	Premature Discharge of Battery	24/12/2021	10/03/2022
							F1905	Device Revision or Replacement	A0705	Battery Problem	24/12/2021	10/03/2022
									A070504	Premature Discharge of Battery	24/12/2021	10/03/2022
					CT981	Single use	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem	24/12/2021	10/03/2022
									A070504	Premature Discharge of Battery	24/12/2021	10/03/2022
							F08	Hospitalization or Prolonged Hospitalization	A0705	Battery Problem	24/12/2021	10/03/2022
									A070504	Premature Discharge of Battery	24/12/2021	10/03/2022
							F1905	Device Revision or Replacement	A0705	Battery Problem	24/12/2021	10/03/2022
									A070504	Premature Discharge of Battery	24/12/2021	10/03/2022
					CT983	Surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem	24/12/2021	10/03/2022
									A070504	Premature Discharge of Battery	24/12/2021	10/03/2022
							F08	Hospitalization or Prolonged Hospitalization	A0705	Battery Problem	24/12/2021	10/03/2022
									A070504	Premature Discharge of Battery	24/12/2021	10/03/2022
							F1905	Device Revision or Replacement	A0705	Battery Problem	24/12/2021	10/03/2022
									A070504	Premature Discharge of Battery	24/12/2021	10/03/2022
7521	2022	All other reportable incidents	35223	Endocardial/interventricular septal pacing lead	CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070909	Oversensing		
									A090801	Signal Artifact/Noise		
							F26	No Health Consequences or Impact	A070909	Oversensing		
									A090801	Signal Artifact/Noise		
7522	2022	All other reportable incidents	47265	Dual-chamber implantable pacemaker, rate-responsive	CT141	Electrodes/Leads and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	29/03/2022	29/03/2022
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	29/03/2022	29/03/2022
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	29/03/2022	29/03/2022
					CT145	Cardiovascular devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	29/03/2022	29/03/2022
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	29/03/2022	29/03/2022
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	29/03/2022	29/03/2022
					CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	29/03/2022	29/03/2022
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	29/03/2022	29/03/2022
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	29/03/2022	29/03/2022

					CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	29/03/2022	29/03/2022
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	29/03/2022	29/03/2022
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	29/03/2022	29/03/2022
					CT233	Surgical	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	29/03/2022	29/03/2022
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	29/03/2022	29/03/2022
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	29/03/2022	29/03/2022
					CT2406	Implantable	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	29/03/2022	29/03/2022
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	29/03/2022	29/03/2022
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	29/03/2022	29/03/2022
					CT2407	Active implantable	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	29/03/2022	29/03/2022
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	29/03/2022	29/03/2022
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	29/03/2022	29/03/2022
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	29/03/2022	29/03/2022
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	29/03/2022	29/03/2022
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	29/03/2022	29/03/2022
					CT2727	Active implant devices/ancillaries	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	29/03/2022	29/03/2022
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	29/03/2022	29/03/2022
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	29/03/2022	29/03/2022
					CT2750	Implantable cardiac electrical stimulators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	29/03/2022	29/03/2022
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	29/03/2022	29/03/2022
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	29/03/2022	29/03/2022
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	29/03/2022	29/03/2022
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	29/03/2022	29/03/2022
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	29/03/2022	29/03/2022
					CT2985	Active	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	29/03/2022	29/03/2022
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	29/03/2022	29/03/2022
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	29/03/2022	29/03/2022
					CT3001	Implantable metal	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	29/03/2022	29/03/2022
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	29/03/2022	29/03/2022
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	29/03/2022	29/03/2022
					CT3002	Implantable synthetic polymer	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	29/03/2022	29/03/2022
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	29/03/2022	29/03/2022
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	29/03/2022	29/03/2022
					CT3006	Implantable inorganic material	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	29/03/2022	29/03/2022
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	29/03/2022	29/03/2022
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	29/03/2022	29/03/2022
					CT314	Electrically powered	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	29/03/2022	29/03/2022
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	29/03/2022	29/03/2022
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	29/03/2022	29/03/2022
					CT321	Long-term surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	29/03/2022	29/03/2022
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	29/03/2022	29/03/2022
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	29/03/2022	29/03/2022
					CT449	Generators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	29/03/2022	29/03/2022
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	29/03/2022	29/03/2022
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	29/03/2022	29/03/2022
					CT981	Single use	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	29/03/2022	29/03/2022
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	29/03/2022	29/03/2022
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	29/03/2022	29/03/2022
					CT983	Surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	29/03/2022	29/03/2022
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	29/03/2022	29/03/2022
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	29/03/2022	29/03/2022
7523	2022	All other reportable incidents	36241	Pacemaker/defibrillator lead	CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A072201	High Impedance	21/12/2021	21/12/2021
							F1908	Prolonged Surgery	A072201	High Impedance	21/12/2021	21/12/2021
							F2203	Imaging Required	A072201	High Impedance	21/12/2021	21/12/2021
7524	2022	All other reportable incidents	35852	Single-chamber implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A0508	Noise, Audible		
									A072201	High Impedance		
							F26	No Health Consequences or Impact	A0508	Noise, Audible		
									A072201	High Impedance		
7525	2022	All other reportable incidents	36241	Pacemaker/defibrillator lead	CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A072201	High Impedance		
							F26	No Health Consequences or Impact	A072201	High Impedance		
7526	2022	All other reportable incidents	17882	Physiologic-monitoring defibrillation system	CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A071301	Failure to Deliver Shock/Stimulation		
							F26	No Health Consequences or Impact	A071301	Failure to Deliver Shock/Stimulation		
7527	2022	All other reportable incidents	35223	Endocardial/interventricular septal pacing lead	CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070909	Oversensing		
									A090801	Signal Artifact/Noise		
							F26	No Health Consequences or Impact	A070909	Oversensing		
									A090801	Signal Artifact/Noise		
7528	2022	All other reportable incidents	35223	Endocardial/interventricular septal pacing lead	CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070909	Oversensing		
									A090801	Signal Artifact/Noise		
							F26	No Health Consequences or Impact	A070909	Oversensing		
									A090801	Signal Artifact/Noise		
7529	2022	All other reportable incidents	36241	Pacemaker/defibrillator lead	CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A072201	High Impedance		
							F26	No Health Consequences or Impact	A072201	High Impedance		
7530	2022	All other reportable incidents	47263	Cardiac resynchronization therapy implantable pacemaker	CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A0908	Incorrect, Inadequate or Imprecise Result or Readings	15/11/2021	15/11/2021
							F08	Hospitalization or Prolonged Hospitalization	A0908	Incorrect, Inadequate or Imprecise Result or Readings	15/11/2021	15/11/2021
							F1905	Device Revision or Replacement	A0908	Incorrect, Inadequate or Imprecise Result or Readings	15/11/2021	15/11/2021
					CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A0908	Incorrect, Inadequate or Imprecise Result or Readings	15/11/2021	15/11/2021
							F08	Hospitalization or Prolonged Hospitalization	A0908	Incorrect, Inadequate or Imprecise Result or Readings	15/11/2021	15/11/2021
							F1905	Device Revision or Replacement	A0908	Incorrect, Inadequate or Imprecise Result or Readings	15/11/2021	15/11/2021
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A0908	Incorrect, Inadequate or Imprecise Result or Readings	15/11/2021	15/11/2021
							F08	Hospitalization or Prolonged Hospitalization	A0908	Incorrect, Inadequate or Imprecise Result or Readings	15/11/2021	15/11/2021
							F1905	Device Revision or Replacement	A0908	Incorrect, Inadequate or Imprecise Result or Readings	15/11/2021	15/11/2021

					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A0908	Incorrect, Inadequate or Imprecise Result or Readings	15/11/2021	15/11/2021
							F08	Hospitalization or Prolonged Hospitalization	A0908	Incorrect, Inadequate or Imprecise Result or Readings	15/11/2021	15/11/2021
							F1905	Device Revision or Replacement	A0908	Incorrect, Inadequate or Imprecise Result or Readings	15/11/2021	15/11/2021
7531	2022	All other reportable incidents	17882	Physiologic-monitoring defibrillation system	CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070803	Failure to Power Up		
							F27	Problem Identified During Non-Clinical Procedure	A070803	Failure to Power Up		
7532	2022	All other reportable incidents	47265	Dual-chamber implantable pacemaker, rate-responsive	CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	29/01/2024	29/01/2024
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	29/01/2024	29/01/2024
							F1903	Device Explantation	A070504	Premature Discharge of Battery	29/01/2024	29/01/2024
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	29/01/2024	29/01/2024
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	29/01/2024	29/01/2024
							F1903	Device Explantation	A070504	Premature Discharge of Battery	29/01/2024	29/01/2024
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	29/01/2024	29/01/2024
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	29/01/2024	29/01/2024
							F1903	Device Explantation	A070504	Premature Discharge of Battery	29/01/2024	29/01/2024
7533	2022	Death	60789	Intracardiac pacemaker	CT130	Catheters and associated devices	E0605	Cardiac Tamponade	A1502	Positioning Problem	28/07/2022	28/07/2022
									A24	No Apparent Device Problem	28/07/2022	28/07/2022
							E0619	Pericardial Effusion	A1502	Positioning Problem	28/07/2022	28/07/2022
									A24	No Apparent Device Problem	28/07/2022	28/07/2022
							E2114	Perforation	A1502	Positioning Problem	28/07/2022	28/07/2022
									A24	No Apparent Device Problem	28/07/2022	28/07/2022
							F02	Death	A1502	Positioning Problem	28/07/2022	28/07/2022
									A24	No Apparent Device Problem	28/07/2022	28/07/2022
					CT145	Cardiovascular devices	E0605	Cardiac Tamponade	A1502	Positioning Problem	28/07/2022	28/07/2022
									A24	No Apparent Device Problem	28/07/2022	28/07/2022
							E0619	Pericardial Effusion	A1502	Positioning Problem	28/07/2022	28/07/2022
									A24	No Apparent Device Problem	28/07/2022	28/07/2022
							E2114	Perforation	A1502	Positioning Problem	28/07/2022	28/07/2022
									A24	No Apparent Device Problem	28/07/2022	28/07/2022
							F02	Death	A1502	Positioning Problem	28/07/2022	28/07/2022
									A24	No Apparent Device Problem	28/07/2022	28/07/2022
					CT1684	Cardiac pacemakers and associated devices	E0605	Cardiac Tamponade	A1502	Positioning Problem	28/07/2022	28/07/2022
									A24	No Apparent Device Problem	28/07/2022	28/07/2022
							E0619	Pericardial Effusion	A1502	Positioning Problem	28/07/2022	28/07/2022
									A24	No Apparent Device Problem	28/07/2022	28/07/2022
							E2114	Perforation	A1502	Positioning Problem	28/07/2022	28/07/2022
									A24	No Apparent Device Problem	28/07/2022	28/07/2022
							F02	Death	A1502	Positioning Problem	28/07/2022	28/07/2022
									A24	No Apparent Device Problem	28/07/2022	28/07/2022
					CT1685	Cardiac pacemakers	E0605	Cardiac Tamponade	A1502	Positioning Problem	28/07/2022	28/07/2022
									A24	No Apparent Device Problem	28/07/2022	28/07/2022
							E0619	Pericardial Effusion	A1502	Positioning Problem	28/07/2022	28/07/2022
									A24	No Apparent Device Problem	28/07/2022	28/07/2022
							E2114	Perforation	A1502	Positioning Problem	28/07/2022	28/07/2022
									A24	No Apparent Device Problem	28/07/2022	28/07/2022
							F02	Death	A1502	Positioning Problem	28/07/2022	28/07/2022
									A24	No Apparent Device Problem	28/07/2022	28/07/2022
					CT233	Surgical	E0605	Cardiac Tamponade	A1502	Positioning Problem	28/07/2022	28/07/2022
									A24	No Apparent Device Problem	28/07/2022	28/07/2022
							E0619	Pericardial Effusion	A1502	Positioning Problem	28/07/2022	28/07/2022
									A24	No Apparent Device Problem	28/07/2022	28/07/2022
							E2114	Perforation	A1502	Positioning Problem	28/07/2022	28/07/2022
									A24	No Apparent Device Problem	28/07/2022	28/07/2022
							F02	Death	A1502	Positioning Problem	28/07/2022	28/07/2022
									A24	No Apparent Device Problem	28/07/2022	28/07/2022
					CT2406	Implantable	E0605	Cardiac Tamponade	A1502	Positioning Problem	28/07/2022	28/07/2022
									A24	No Apparent Device Problem	28/07/2022	28/07/2022
							E0619	Pericardial Effusion	A1502	Positioning Problem	28/07/2022	28/07/2022
									A24	No Apparent Device Problem	28/07/2022	28/07/2022
							E2114	Perforation	A1502	Positioning Problem	28/07/2022	28/07/2022
									A24	No Apparent Device Problem	28/07/2022	28/07/2022
							F02	Death	A1502	Positioning Problem	28/07/2022	28/07/2022
									A24	No Apparent Device Problem	28/07/2022	28/07/2022
					CT2407	Active implantable	E0605	Cardiac Tamponade	A1502	Positioning Problem	28/07/2022	28/07/2022
									A24	No Apparent Device Problem	28/07/2022	28/07/2022
							E0619	Pericardial Effusion	A1502	Positioning Problem	28/07/2022	28/07/2022
									A24	No Apparent Device Problem	28/07/2022	28/07/2022
							E2114	Perforation	A1502	Positioning Problem	28/07/2022	28/07/2022
									A24	No Apparent Device Problem	28/07/2022	28/07/2022
							F02	Death	A1502	Positioning Problem	28/07/2022	28/07/2022
									A24	No Apparent Device Problem	28/07/2022	28/07/2022
					CT257	Implantable cardiac pacemakers	E0605	Cardiac Tamponade	A1502	Positioning Problem	28/07/2022	28/07/2022
									A24	No Apparent Device Problem	28/07/2022	28/07/2022
							E0619	Pericardial Effusion	A1502	Positioning Problem	28/07/2022	28/07/2022
									A24	No Apparent Device Problem	28/07/2022	28/07/2022
							E2114	Perforation	A1502	Positioning Problem	28/07/2022	28/07/2022
									A24	No Apparent Device Problem	28/07/2022	28/07/2022
							F02	Death	A1502	Positioning Problem	28/07/2022	28/07/2022
									A24	No Apparent Device Problem	28/07/2022	28/07/2022
					CT2727	Active implant devices/ancillaries	E0605	Cardiac Tamponade	A1502	Positioning Problem	28/07/2022	28/07/2022
									A24	No Apparent Device Problem	28/07/2022	28/07/2022
							E0619	Pericardial Effusion	A1502	Positioning Problem	28/07/2022	28/07/2022
									A24	No Apparent Device Problem	28/07/2022	28/07/2022
							E2114	Perforation	A1502	Positioning Problem	28/07/2022	28/07/2022
									A24	No Apparent Device Problem	28/07/2022	28/07/2022
							F02	Death	A1502	Positioning Problem	28/07/2022	28/07/2022

									A24	No Apparent Device Problem	28/07/2022	28/07/2022
					CT2750	Implantable cardiac electrical stimulators and associated devices	E0605	Cardiac Tamponade	A1502	Positioning Problem	28/07/2022	28/07/2022
									A24	No Apparent Device Problem	28/07/2022	28/07/2022
							E0619	Pericardial Effusion	A1502	Positioning Problem	28/07/2022	28/07/2022
									A24	No Apparent Device Problem	28/07/2022	28/07/2022
							E2114	Perforation	A1502	Positioning Problem	28/07/2022	28/07/2022
									A24	No Apparent Device Problem	28/07/2022	28/07/2022
							F02	Death	A1502	Positioning Problem	28/07/2022	28/07/2022
									A24	No Apparent Device Problem	28/07/2022	28/07/2022
					CT2818	Catheterization/Cannulation/Fluid path devices	E0605	Cardiac Tamponade	A1502	Positioning Problem	28/07/2022	28/07/2022
									A24	No Apparent Device Problem	28/07/2022	28/07/2022
							E0619	Pericardial Effusion	A1502	Positioning Problem	28/07/2022	28/07/2022
									A24	No Apparent Device Problem	28/07/2022	28/07/2022
							E2114	Perforation	A1502	Positioning Problem	28/07/2022	28/07/2022
									A24	No Apparent Device Problem	28/07/2022	28/07/2022
							F02	Death	A1502	Positioning Problem	28/07/2022	28/07/2022
									A24	No Apparent Device Problem	28/07/2022	28/07/2022
					CT2866	Implantable cardiac electrical stimulators	E0605	Cardiac Tamponade	A1502	Positioning Problem	28/07/2022	28/07/2022
									A24	No Apparent Device Problem	28/07/2022	28/07/2022
							E0619	Pericardial Effusion	A1502	Positioning Problem	28/07/2022	28/07/2022
									A24	No Apparent Device Problem	28/07/2022	28/07/2022
							E2114	Perforation	A1502	Positioning Problem	28/07/2022	28/07/2022
									A24	No Apparent Device Problem	28/07/2022	28/07/2022
							F02	Death	A1502	Positioning Problem	28/07/2022	28/07/2022
									A24	No Apparent Device Problem	28/07/2022	28/07/2022
					CT2985	Active	E0605	Cardiac Tamponade	A1502	Positioning Problem	28/07/2022	28/07/2022
									A24	No Apparent Device Problem	28/07/2022	28/07/2022
							E0619	Pericardial Effusion	A1502	Positioning Problem	28/07/2022	28/07/2022
									A24	No Apparent Device Problem	28/07/2022	28/07/2022
							E2114	Perforation	A1502	Positioning Problem	28/07/2022	28/07/2022
									A24	No Apparent Device Problem	28/07/2022	28/07/2022
							F02	Death	A1502	Positioning Problem	28/07/2022	28/07/2022
									A24	No Apparent Device Problem	28/07/2022	28/07/2022
					CT3001	Implantable metal	E0605	Cardiac Tamponade	A1502	Positioning Problem	28/07/2022	28/07/2022
									A24	No Apparent Device Problem	28/07/2022	28/07/2022
							E0619	Pericardial Effusion	A1502	Positioning Problem	28/07/2022	28/07/2022
									A24	No Apparent Device Problem	28/07/2022	28/07/2022
							E2114	Perforation	A1502	Positioning Problem	28/07/2022	28/07/2022
									A24	No Apparent Device Problem	28/07/2022	28/07/2022
							F02	Death	A1502	Positioning Problem	28/07/2022	28/07/2022
									A24	No Apparent Device Problem	28/07/2022	28/07/2022
					CT3002	Implantable synthetic polymer	E0605	Cardiac Tamponade	A1502	Positioning Problem	28/07/2022	28/07/2022
									A24	No Apparent Device Problem	28/07/2022	28/07/2022
							E0619	Pericardial Effusion	A1502	Positioning Problem	28/07/2022	28/07/2022
									A24	No Apparent Device Problem	28/07/2022	28/07/2022
							E2114	Perforation	A1502	Positioning Problem	28/07/2022	28/07/2022
									A24	No Apparent Device Problem	28/07/2022	28/07/2022
							F02	Death	A1502	Positioning Problem	28/07/2022	28/07/2022
									A24	No Apparent Device Problem	28/07/2022	28/07/2022
					CT3006	Implantable inorganic material	E0605	Cardiac Tamponade	A1502	Positioning Problem	28/07/2022	28/07/2022
									A24	No Apparent Device Problem	28/07/2022	28/07/2022
							E0619	Pericardial Effusion	A1502	Positioning Problem	28/07/2022	28/07/2022
									A24	No Apparent Device Problem	28/07/2022	28/07/2022
							E2114	Perforation	A1502	Positioning Problem	28/07/2022	28/07/2022
									A24	No Apparent Device Problem	28/07/2022	28/07/2022
							F02	Death	A1502	Positioning Problem	28/07/2022	28/07/2022
									A24	No Apparent Device Problem	28/07/2022	28/07/2022
					CT314	Electrically powered	E0605	Cardiac Tamponade	A1502	Positioning Problem	28/07/2022	28/07/2022
									A24	No Apparent Device Problem	28/07/2022	28/07/2022
							E0619	Pericardial Effusion	A1502	Positioning Problem	28/07/2022	28/07/2022
									A24	No Apparent Device Problem	28/07/2022	28/07/2022
							E2114	Perforation	A1502	Positioning Problem	28/07/2022	28/07/2022
									A24	No Apparent Device Problem	28/07/2022	28/07/2022
							F02	Death	A1502	Positioning Problem	28/07/2022	28/07/2022
									A24	No Apparent Device Problem	28/07/2022	28/07/2022
					CT321	Long-term surgical invasive	E0605	Cardiac Tamponade	A1502	Positioning Problem	28/07/2022	28/07/2022
									A24	No Apparent Device Problem	28/07/2022	28/07/2022
							E0619	Pericardial Effusion	A1502	Positioning Problem	28/07/2022	28/07/2022
									A24	No Apparent Device Problem	28/07/2022	28/07/2022
							E2114	Perforation	A1502	Positioning Problem	28/07/2022	28/07/2022
									A24	No Apparent Device Problem	28/07/2022	28/07/2022
							F02	Death	A1502	Positioning Problem	28/07/2022	28/07/2022
									A24	No Apparent Device Problem	28/07/2022	28/07/2022
					CT981	Single use	E0605	Cardiac Tamponade	A1502	Positioning Problem	28/07/2022	28/07/2022
									A24	No Apparent Device Problem	28/07/2022	28/07/2022
							E0619	Pericardial Effusion	A1502	Positioning Problem	28/07/2022	28/07/2022
									A24	No Apparent Device Problem	28/07/2022	28/07/2022
							E2114	Perforation	A1502	Positioning Problem	28/07/2022	28/07/2022
									A24	No Apparent Device Problem	28/07/2022	28/07/2022
							F02	Death	A1502	Positioning Problem	28/07/2022	28/07/2022
									A24	No Apparent Device Problem	28/07/2022	28/07/2022
					CT983	Surgical invasive	E0605	Cardiac Tamponade	A1502	Positioning Problem	28/07/2022	28/07/2022
									A24	No Apparent Device Problem	28/07/2022	28/07/2022
							E0619	Pericardial Effusion	A1502	Positioning Problem	28/07/2022	28/07/2022
									A24	No Apparent Device Problem	28/07/2022	28/07/2022
							E2114	Perforation	A1502	Positioning Problem	28/07/2022	28/07/2022
									A24	No Apparent Device Problem	28/07/2022	28/07/2022
							F02	Death	A1502	Positioning Problem	28/07/2022	28/07/2022

									A24	No Apparent Device Problem	28/07/2022	28/07/2022
7534	2022	All other reportable incidents	17882	Physiologic-monitoring defibrillation system	CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A0908	Incorrect, Inadequate or Imprecise Result or Readings		
							F27	Problem Identified During Non-Clinical Procedure	A0908	Incorrect, Inadequate or Imprecise Result or Readings		
7535	2022	All other reportable incidents	36241	Pacemaker/defibrillator lead	CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A072201	High Impedance	18/08/2022	18/08/2022
							F08	Hospitalization or Prolonged Hospitalization	A072201	High Impedance	18/08/2022	18/08/2022
							F1905	Device Revision or Replacement	A072201	High Impedance	18/08/2022	18/08/2022
7536	2022	All other reportable incidents	47265	Dual-chamber implantable pacemaker, rate-responsive	CT141	Electrodes/Leads and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem	02/11/2022	02/11/2022
									A070504	Premature Discharge of Battery	02/11/2022	02/11/2022
							F08	Hospitalization or Prolonged Hospitalization	A0705	Battery Problem	02/11/2022	02/11/2022
									A070504	Premature Discharge of Battery	02/11/2022	02/11/2022
							F1905	Device Revision or Replacement	A0705	Battery Problem	02/11/2022	02/11/2022
									A070504	Premature Discharge of Battery	02/11/2022	02/11/2022
					CT145	Cardiovascular devices	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem	02/11/2022	02/11/2022
									A070504	Premature Discharge of Battery	02/11/2022	02/11/2022
							F08	Hospitalization or Prolonged Hospitalization	A0705	Battery Problem	02/11/2022	02/11/2022
									A070504	Premature Discharge of Battery	02/11/2022	02/11/2022
							F1905	Device Revision or Replacement	A0705	Battery Problem	02/11/2022	02/11/2022
									A070504	Premature Discharge of Battery	02/11/2022	02/11/2022
					CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem	02/11/2022	02/11/2022
									A070504	Premature Discharge of Battery	02/11/2022	02/11/2022
							F08	Hospitalization or Prolonged Hospitalization	A0705	Battery Problem	02/11/2022	02/11/2022
									A070504	Premature Discharge of Battery	02/11/2022	02/11/2022
							F1905	Device Revision or Replacement	A0705	Battery Problem	02/11/2022	02/11/2022
									A070504	Premature Discharge of Battery	02/11/2022	02/11/2022
					CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem	02/11/2022	02/11/2022
									A070504	Premature Discharge of Battery	02/11/2022	02/11/2022
							F08	Hospitalization or Prolonged Hospitalization	A0705	Battery Problem	02/11/2022	02/11/2022
									A070504	Premature Discharge of Battery	02/11/2022	02/11/2022
							F1905	Device Revision or Replacement	A0705	Battery Problem	02/11/2022	02/11/2022
									A070504	Premature Discharge of Battery	02/11/2022	02/11/2022
					CT233	Surgical	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem	02/11/2022	02/11/2022
									A070504	Premature Discharge of Battery	02/11/2022	02/11/2022
							F08	Hospitalization or Prolonged Hospitalization	A0705	Battery Problem	02/11/2022	02/11/2022
									A070504	Premature Discharge of Battery	02/11/2022	02/11/2022
							F1905	Device Revision or Replacement	A0705	Battery Problem	02/11/2022	02/11/2022
									A070504	Premature Discharge of Battery	02/11/2022	02/11/2022
					CT2406	Implantable	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem	02/11/2022	02/11/2022
									A070504	Premature Discharge of Battery	02/11/2022	02/11/2022
							F08	Hospitalization or Prolonged Hospitalization	A0705	Battery Problem	02/11/2022	02/11/2022
									A070504	Premature Discharge of Battery	02/11/2022	02/11/2022
							F1905	Device Revision or Replacement	A0705	Battery Problem	02/11/2022	02/11/2022
									A070504	Premature Discharge of Battery	02/11/2022	02/11/2022
					CT2407	Active implantable	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem	02/11/2022	02/11/2022
									A070504	Premature Discharge of Battery	02/11/2022	02/11/2022
							F08	Hospitalization or Prolonged Hospitalization	A0705	Battery Problem	02/11/2022	02/11/2022
									A070504	Premature Discharge of Battery	02/11/2022	02/11/2022
							F1905	Device Revision or Replacement	A0705	Battery Problem	02/11/2022	02/11/2022
									A070504	Premature Discharge of Battery	02/11/2022	02/11/2022
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem	02/11/2022	02/11/2022
									A070504	Premature Discharge of Battery	02/11/2022	02/11/2022
							F08	Hospitalization or Prolonged Hospitalization	A0705	Battery Problem	02/11/2022	02/11/2022
									A070504	Premature Discharge of Battery	02/11/2022	02/11/2022
							F1905	Device Revision or Replacement	A0705	Battery Problem	02/11/2022	02/11/2022
									A070504	Premature Discharge of Battery	02/11/2022	02/11/2022
					CT2727	Active implant devices/ancillaries	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem	02/11/2022	02/11/2022
									A070504	Premature Discharge of Battery	02/11/2022	02/11/2022
							F08	Hospitalization or Prolonged Hospitalization	A0705	Battery Problem	02/11/2022	02/11/2022
									A070504	Premature Discharge of Battery	02/11/2022	02/11/2022
							F1905	Device Revision or Replacement	A0705	Battery Problem	02/11/2022	02/11/2022
									A070504	Premature Discharge of Battery	02/11/2022	02/11/2022
					CT2750	Implantable cardiac electrical stimulators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem	02/11/2022	02/11/2022
									A070504	Premature Discharge of Battery	02/11/2022	02/11/2022
							F08	Hospitalization or Prolonged Hospitalization	A0705	Battery Problem	02/11/2022	02/11/2022
									A070504	Premature Discharge of Battery	02/11/2022	02/11/2022
							F1905	Device Revision or Replacement	A0705	Battery Problem	02/11/2022	02/11/2022
									A070504	Premature Discharge of Battery	02/11/2022	02/11/2022
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem	02/11/2022	02/11/2022
									A070504	Premature Discharge of Battery	02/11/2022	02/11/2022
							F08	Hospitalization or Prolonged Hospitalization	A0705	Battery Problem	02/11/2022	02/11/2022
									A070504	Premature Discharge of Battery	02/11/2022	02/11/2022
							F1905	Device Revision or Replacement	A0705	Battery Problem	02/11/2022	02/11/2022
									A070504	Premature Discharge of Battery	02/11/2022	02/11/2022
					CT2985	Active	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem	02/11/2022	02/11/2022
									A070504	Premature Discharge of Battery	02/11/2022	02/11/2022
							F08	Hospitalization or Prolonged Hospitalization	A0705	Battery Problem	02/11/2022	02/11/2022
									A070504	Premature Discharge of Battery	02/11/2022	02/11/2022
							F1905	Device Revision or Replacement	A0705	Battery Problem	02/11/2022	02/11/2022
									A070504	Premature Discharge of Battery	02/11/2022	02/11/2022
					CT3001	Implantable metal	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem	02/11/2022	02/11/2022
									A070504	Premature Discharge of Battery	02/11/2022	02/11/2022
							F08	Hospitalization or Prolonged Hospitalization	A0705	Battery Problem	02/11/2022	02/11/2022
									A070504	Premature Discharge of Battery	02/11/2022	02/11/2022
							F1905	Device Revision or Replacement	A0705	Battery Problem	02/11/2022	02/11/2022
									A070504	Premature Discharge of Battery	02/11/2022	02/11/2022
					CT3002	Implantable synthetic polymer	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem	02/11/2022	02/11/2022
									A070504	Premature Discharge of Battery	02/11/2022	02/11/2022

							F08	Hospitalization or Prolonged Hospitalization	A0705	Battery Problem	02/11/2022	02/11/2022
									A070504	Premature Discharge of Battery	02/11/2022	02/11/2022
							F1905	Device Revision or Replacement	A0705	Battery Problem	02/11/2022	02/11/2022
									A070504	Premature Discharge of Battery	02/11/2022	02/11/2022
					CT3006	Implantable inorganic material	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem	02/11/2022	02/11/2022
									A070504	Premature Discharge of Battery	02/11/2022	02/11/2022
							F08	Hospitalization or Prolonged Hospitalization	A0705	Battery Problem	02/11/2022	02/11/2022
									A070504	Premature Discharge of Battery	02/11/2022	02/11/2022
							F1905	Device Revision or Replacement	A0705	Battery Problem	02/11/2022	02/11/2022
									A070504	Premature Discharge of Battery	02/11/2022	02/11/2022
					CT314	Electrically powered	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem	02/11/2022	02/11/2022
									A070504	Premature Discharge of Battery	02/11/2022	02/11/2022
							F08	Hospitalization or Prolonged Hospitalization	A0705	Battery Problem	02/11/2022	02/11/2022
									A070504	Premature Discharge of Battery	02/11/2022	02/11/2022
							F1905	Device Revision or Replacement	A0705	Battery Problem	02/11/2022	02/11/2022
									A070504	Premature Discharge of Battery	02/11/2022	02/11/2022
					CT321	Long-term surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem	02/11/2022	02/11/2022
									A070504	Premature Discharge of Battery	02/11/2022	02/11/2022
							F08	Hospitalization or Prolonged Hospitalization	A0705	Battery Problem	02/11/2022	02/11/2022
									A070504	Premature Discharge of Battery	02/11/2022	02/11/2022
							F1905	Device Revision or Replacement	A0705	Battery Problem	02/11/2022	02/11/2022
									A070504	Premature Discharge of Battery	02/11/2022	02/11/2022
					CT449	Generators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem	02/11/2022	02/11/2022
									A070504	Premature Discharge of Battery	02/11/2022	02/11/2022
							F08	Hospitalization or Prolonged Hospitalization	A0705	Battery Problem	02/11/2022	02/11/2022
									A070504	Premature Discharge of Battery	02/11/2022	02/11/2022
							F1905	Device Revision or Replacement	A0705	Battery Problem	02/11/2022	02/11/2022
									A070504	Premature Discharge of Battery	02/11/2022	02/11/2022
					CT981	Single use	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem	02/11/2022	02/11/2022
									A070504	Premature Discharge of Battery	02/11/2022	02/11/2022
							F08	Hospitalization or Prolonged Hospitalization	A0705	Battery Problem	02/11/2022	02/11/2022
									A070504	Premature Discharge of Battery	02/11/2022	02/11/2022
							F1905	Device Revision or Replacement	A0705	Battery Problem	02/11/2022	02/11/2022
									A070504	Premature Discharge of Battery	02/11/2022	02/11/2022
					CT983	Surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem	02/11/2022	02/11/2022
									A070504	Premature Discharge of Battery	02/11/2022	02/11/2022
							F08	Hospitalization or Prolonged Hospitalization	A0705	Battery Problem	02/11/2022	02/11/2022
									A070504	Premature Discharge of Battery	02/11/2022	02/11/2022
							F1905	Device Revision or Replacement	A0705	Battery Problem	02/11/2022	02/11/2022
									A070504	Premature Discharge of Battery	02/11/2022	02/11/2022
7537	2022	All other reportable incidents	63188	Left-ventricular wireless pacing system sensor/transmitter	CT141	Electrodes/Leads and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A010102	Device Appears to Trigger Rejection	13/06/2022	13/06/2022
					CT145	Cardiovascular devices	E2403	No Clinical Signs, Symptoms or Conditions	A010102	Device Appears to Trigger Rejection	13/06/2022	13/06/2022
					CT1577	Wireless monitoring system transmitters	E2403	No Clinical Signs, Symptoms or Conditions	A010102	Device Appears to Trigger Rejection	13/06/2022	13/06/2022
					CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A010102	Device Appears to Trigger Rejection	13/06/2022	13/06/2022
					CT2317	Transmitters/Receivers/Transceivers	E2403	No Clinical Signs, Symptoms or Conditions	A010102	Device Appears to Trigger Rejection	13/06/2022	13/06/2022
					CT233	Surgical	E2403	No Clinical Signs, Symptoms or Conditions	A010102	Device Appears to Trigger Rejection	13/06/2022	13/06/2022
					CT2406	Implantable	E2403	No Clinical Signs, Symptoms or Conditions	A010102	Device Appears to Trigger Rejection	13/06/2022	13/06/2022
					CT2407	Active implantable	E2403	No Clinical Signs, Symptoms or Conditions	A010102	Device Appears to Trigger Rejection	13/06/2022	13/06/2022
					CT255	Transducers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A010102	Device Appears to Trigger Rejection	13/06/2022	13/06/2022
					CT2726	Physiological monitoring/measuring devices	E2403	No Clinical Signs, Symptoms or Conditions	A010102	Device Appears to Trigger Rejection	13/06/2022	13/06/2022
					CT2727	Active implant devices/ancillaries	E2403	No Clinical Signs, Symptoms or Conditions	A010102	Device Appears to Trigger Rejection	13/06/2022	13/06/2022
					CT2750	Implantable cardiac electrical stimulators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A010102	Device Appears to Trigger Rejection	13/06/2022	13/06/2022
					CT2885	Physiological monitoring/measuring system transmitters/receivers/transceivers	E2403	No Clinical Signs, Symptoms or Conditions	A010102	Device Appears to Trigger Rejection	13/06/2022	13/06/2022
					CT2985	Active	E2403	No Clinical Signs, Symptoms or Conditions	A010102	Device Appears to Trigger Rejection	13/06/2022	13/06/2022
					CT3001	Implantable metal	E2403	No Clinical Signs, Symptoms or Conditions	A010102	Device Appears to Trigger Rejection	13/06/2022	13/06/2022
					CT3002	Implantable synthetic polymer	E2403	No Clinical Signs, Symptoms or Conditions	A010102	Device Appears to Trigger Rejection	13/06/2022	13/06/2022
					CT3006	Implantable inorganic material	E2403	No Clinical Signs, Symptoms or Conditions	A010102	Device Appears to Trigger Rejection	13/06/2022	13/06/2022
					CT314	Electrically powered	E2403	No Clinical Signs, Symptoms or Conditions	A010102	Device Appears to Trigger Rejection	13/06/2022	13/06/2022
					CT321	Long-term surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A010102	Device Appears to Trigger Rejection	13/06/2022	13/06/2022
					CT384	Sensors and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A010102	Device Appears to Trigger Rejection	13/06/2022	13/06/2022
					CT449	Generators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A010102	Device Appears to Trigger Rejection	13/06/2022	13/06/2022
					CT981	Single use	E2403	No Clinical Signs, Symptoms or Conditions	A010102	Device Appears to Trigger Rejection	13/06/2022	13/06/2022
					CT983	Surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A010102	Device Appears to Trigger Rejection	13/06/2022	13/06/2022
					CT994	Sensors	E2403	No Clinical Signs, Symptoms or Conditions	A010102	Device Appears to Trigger Rejection	13/06/2022	13/06/2022
7538	2022	All other reportable incidents	47106	Drug-eluting endocardial pacing lead	CT1684	Cardiac pacemakers and associated devices	E2401	Insufficient Information	A070102	High Capture Threshold	16/08/2022	
							E2403	No Clinical Signs, Symptoms or Conditions	A070102	High Capture Threshold	16/08/2022	
							F0101	Therapeutic Response Decreased	A070102	High Capture Threshold	16/08/2022	
7539	2022	All other reportable incidents	17882	Physiologic-monitoring defibrillation system	CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070603	Failure to Charge		
							F27	Problem Identified During Non-Clinical Procedure	A070603	Failure to Charge		
7540	2022	All other reportable incidents	35223	Endocardial/interventricular septal pacing lead	CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070909	Oversensing		
									A072201	High Impedance		
									A090801	Signal Artifact/Noise		
							F26	No Health Consequences or Impact	A070909	Oversensing		
									A072201	High Impedance		
									A090801	Signal Artifact/Noise		
7541	2022	All other reportable incidents	47265	Dual-chamber implantable pacemaker, rate-responsive	CT141	Electrodes/Leads and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	05/09/2022	05/09/2022
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	05/09/2022	05/09/2022
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	05/09/2022	05/09/2022
					CT145	Cardiovascular devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	05/09/2022	05/09/2022
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	05/09/2022	05/09/2022
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	05/09/2022	05/09/2022
					CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	05/09/2022	05/09/2022
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	05/09/2022	05/09/2022
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	05/09/2022	05/09/2022
					CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	05/09/2022	05/09/2022
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	05/09/2022	05/09/2022

							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	05/09/2022	05/09/2022
					CT233	Surgical	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	05/09/2022	05/09/2022
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	05/09/2022	05/09/2022
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	05/09/2022	05/09/2022
					CT2406	Implantable	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	05/09/2022	05/09/2022
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	05/09/2022	05/09/2022
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	05/09/2022	05/09/2022
					CT2407	Active implantable	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	05/09/2022	05/09/2022
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	05/09/2022	05/09/2022
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	05/09/2022	05/09/2022
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	05/09/2022	05/09/2022
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	05/09/2022	05/09/2022
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	05/09/2022	05/09/2022
					CT2727	Active implant devices/ancillaries	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	05/09/2022	05/09/2022
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	05/09/2022	05/09/2022
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	05/09/2022	05/09/2022
					CT2750	Implantable cardiac electrical stimulators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	05/09/2022	05/09/2022
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	05/09/2022	05/09/2022
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	05/09/2022	05/09/2022
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	05/09/2022	05/09/2022
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	05/09/2022	05/09/2022
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	05/09/2022	05/09/2022
					CT2985	Active	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	05/09/2022	05/09/2022
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	05/09/2022	05/09/2022
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	05/09/2022	05/09/2022
					CT3001	Implantable metal	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	05/09/2022	05/09/2022
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	05/09/2022	05/09/2022
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	05/09/2022	05/09/2022
					CT3002	Implantable synthetic polymer	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	05/09/2022	05/09/2022
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	05/09/2022	05/09/2022
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	05/09/2022	05/09/2022
					CT3006	Implantable inorganic material	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	05/09/2022	05/09/2022
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	05/09/2022	05/09/2022
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	05/09/2022	05/09/2022
					CT314	Electrically powered	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	05/09/2022	05/09/2022
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	05/09/2022	05/09/2022
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	05/09/2022	05/09/2022
					CT321	Long-term surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	05/09/2022	05/09/2022
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	05/09/2022	05/09/2022
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	05/09/2022	05/09/2022
					CT449	Generators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	05/09/2022	05/09/2022
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	05/09/2022	05/09/2022
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	05/09/2022	05/09/2022
					CT981	Single use	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	05/09/2022	05/09/2022
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	05/09/2022	05/09/2022
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	05/09/2022	05/09/2022
					CT983	Surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	05/09/2022	05/09/2022
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	05/09/2022	05/09/2022
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	05/09/2022	05/09/2022
7542	2022	All other reportable incidents	17882	Physiologic-monitoring defibrillation system	CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A090811	Unable to Obtain Readings		
							F27	Problem Identified During Non-Clinical Procedure	A090811	Unable to Obtain Readings		
7543	2022	All other reportable incidents	17882	Physiologic-monitoring defibrillation system	CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A0721	Circuit Failure		
							F27	Problem Identified During Non-Clinical Procedure	A0721	Circuit Failure		
7544	2022	All other reportable incidents	35852	Single-chamber implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	27/09/2022	27/09/2022
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	27/09/2022	27/09/2022
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	27/09/2022	27/09/2022
7545	2022	All other reportable incidents	35223	Endocardial/interventricular septal pacing lead	CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A0401	Break		
									A040101	Fracture		
									A07	Electrical /Electronic Property Problem		
							F26	No Health Consequences or Impact	A0401	Break		
									A040101	Fracture		
									A07	Electrical /Electronic Property Problem		
7546	2022	All other reportable incidents	47265	Dual-chamber implantable pacemaker, rate-responsive	CT141	Electrodes/Leads and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	17/01/2022	17/01/2022
							F1901	Additional Surgery	A070504	Premature Discharge of Battery	17/01/2022	17/01/2022
					CT145	Cardiovascular devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	17/01/2022	17/01/2022
							F1901	Additional Surgery	A070504	Premature Discharge of Battery	17/01/2022	17/01/2022
					CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	17/01/2022	17/01/2022
							F1901	Additional Surgery	A070504	Premature Discharge of Battery	17/01/2022	17/01/2022
					CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	17/01/2022	17/01/2022
							F1901	Additional Surgery	A070504	Premature Discharge of Battery	17/01/2022	17/01/2022
					CT233	Surgical	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	17/01/2022	17/01/2022
							F1901	Additional Surgery	A070504	Premature Discharge of Battery	17/01/2022	17/01/2022
					CT2406	Implantable	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	17/01/2022	17/01/2022
							F1901	Additional Surgery	A070504	Premature Discharge of Battery	17/01/2022	17/01/2022
					CT2407	Active implantable	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	17/01/2022	17/01/2022
							F1901	Additional Surgery	A070504	Premature Discharge of Battery	17/01/2022	17/01/2022
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	17/01/2022	17/01/2022
							F1901	Additional Surgery	A070504	Premature Discharge of Battery	17/01/2022	17/01/2022
					CT2727	Active implant devices/ancillaries	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	17/01/2022	17/01/2022
							F1901	Additional Surgery	A070504	Premature Discharge of Battery	17/01/2022	17/01/2022
					CT2750	Implantable cardiac electrical stimulators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	17/01/2022	17/01/2022
							F1901	Additional Surgery	A070504	Premature Discharge of Battery	17/01/2022	17/01/2022
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	17/01/2022	17/01/2022
							F1901	Additional Surgery	A070504	Premature Discharge of Battery	17/01/2022	17/01/2022
					CT2985	Active	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	17/01/2022	17/01/2022
							F1901	Additional Surgery	A070504	Premature Discharge of Battery	17/01/2022	17/01/2022
					CT3001	Implantable metal	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	17/01/2022	17/01/2022
							F1901	Additional Surgery	A070504	Premature Discharge of Battery	17/01/2022	17/01/2022

					CT3002	Implantable synthetic polymer	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	17/01/2022	17/01/2022
							F1901	Additional Surgery	A070504	Premature Discharge of Battery	17/01/2022	17/01/2022
					CT3006	Implantable inorganic material	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	17/01/2022	17/01/2022
							F1901	Additional Surgery	A070504	Premature Discharge of Battery	17/01/2022	17/01/2022
					CT314	Electrically powered	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	17/01/2022	17/01/2022
							F1901	Additional Surgery	A070504	Premature Discharge of Battery	17/01/2022	17/01/2022
					CT321	Long-term surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	17/01/2022	17/01/2022
							F1901	Additional Surgery	A070504	Premature Discharge of Battery	17/01/2022	17/01/2022
					CT449	Generators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	17/01/2022	17/01/2022
							F1901	Additional Surgery	A070504	Premature Discharge of Battery	17/01/2022	17/01/2022
					CT981	Single use	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	17/01/2022	17/01/2022
							F1901	Additional Surgery	A070504	Premature Discharge of Battery	17/01/2022	17/01/2022
					CT983	Surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	17/01/2022	17/01/2022
							F1901	Additional Surgery	A070504	Premature Discharge of Battery	17/01/2022	17/01/2022
7547	2022	All other reportable incidents	35852	Single-chamber implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem	21/04/2022	21/04/2022
									A070504	Premature Discharge of Battery	21/04/2022	21/04/2022
							F08	Hospitalization or Prolonged Hospitalization	A0705	Battery Problem	21/04/2022	21/04/2022
									A070504	Premature Discharge of Battery	21/04/2022	21/04/2022
							F1905	Device Revision or Replacement	A0705	Battery Problem	21/04/2022	21/04/2022
									A070504	Premature Discharge of Battery	21/04/2022	21/04/2022
7548	2022	All other reportable incidents	35852	Single-chamber implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem	08/02/2022	08/02/2022
									A070504	Premature Discharge of Battery	08/02/2022	08/02/2022
							F08	Hospitalization or Prolonged Hospitalization	A0705	Battery Problem	08/02/2022	08/02/2022
									A070504	Premature Discharge of Battery	08/02/2022	08/02/2022
							F1905	Device Revision or Replacement	A0705	Battery Problem	08/02/2022	08/02/2022
									A070504	Premature Discharge of Battery	08/02/2022	08/02/2022
7549	2022	All other reportable incidents	35852	Single-chamber implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E2104	Electric Shock	A070909	Oversensing		
									A0712	Pacing Problem		
									A071302	Inappropriate/Inadequate Shock/Stimulation		
							E2403	No Clinical Signs, Symptoms or Conditions	A070909	Oversensing		
									A0712	Pacing Problem		
									A071302	Inappropriate/Inadequate Shock/Stimulation		
							F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A070909	Oversensing		
									A0712	Pacing Problem		
									A071302	Inappropriate/Inadequate Shock/Stimulation		
							F2303	Medication Required	A070909	Oversensing		
									A0712	Pacing Problem		
									A071302	Inappropriate/Inadequate Shock/Stimulation		
7550	2022	All other reportable incidents	35223	Endocardial/interventricular septal pacing lead	CT1684	Cardiac pacemakers and associated devices	E060106	Heart Block	A070909	Oversensing		
									A0722	Impedance Problem		
									A090801	Signal Artifact/Noise		
							E0717	Dyspnea	A070909	Oversensing		
									A0722	Impedance Problem		
									A090801	Signal Artifact/Noise		
							E233801	Peripheral Edema	A070909	Oversensing		
									A0722	Impedance Problem		
									A090801	Signal Artifact/Noise		
							E2403	No Clinical Signs, Symptoms or Conditions	A070909	Oversensing		
									A0722	Impedance Problem		
									A090801	Signal Artifact/Noise		
							F11	Non-Serious Injury/ Illness/ Impairment	A070909	Oversensing		
									A0722	Impedance Problem		
									A090801	Signal Artifact/Noise		
							F23	Unexpected Medical Intervention	A070909	Oversensing		
									A0722	Impedance Problem		
									A090801	Signal Artifact/Noise		
7551	2022	All other reportable incidents	17882	Physiologic-monitoring defibrillation system	CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A0801	Failure to Calibrate		
							F27	Problem Identified During Non-Clinical Procedure	A0801	Failure to Calibrate		
7552	2022	All other reportable incidents	47263	Cardiac resynchronization therapy implantable pacemaker	CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A0908	Incorrect, Inadequate or Imprecise Result or Readings	26/07/2022	26/07/2022
							F08	Hospitalization or Prolonged Hospitalization	A0908	Incorrect, Inadequate or Imprecise Result or Readings	26/07/2022	26/07/2022
							F1905	Device Revision or Replacement	A0908	Incorrect, Inadequate or Imprecise Result or Readings	26/07/2022	26/07/2022
					CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A0908	Incorrect, Inadequate or Imprecise Result or Readings	26/07/2022	26/07/2022
							F08	Hospitalization or Prolonged Hospitalization	A0908	Incorrect, Inadequate or Imprecise Result or Readings	26/07/2022	26/07/2022
							F1905	Device Revision or Replacement	A0908	Incorrect, Inadequate or Imprecise Result or Readings	26/07/2022	26/07/2022
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A0908	Incorrect, Inadequate or Imprecise Result or Readings	26/07/2022	26/07/2022
							F08	Hospitalization or Prolonged Hospitalization	A0908	Incorrect, Inadequate or Imprecise Result or Readings	26/07/2022	26/07/2022
							F1905	Device Revision or Replacement	A0908	Incorrect, Inadequate or Imprecise Result or Readings	26/07/2022	26/07/2022
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A0908	Incorrect, Inadequate or Imprecise Result or Readings	26/07/2022	26/07/2022
							F08	Hospitalization or Prolonged Hospitalization	A0908	Incorrect, Inadequate or Imprecise Result or Readings	26/07/2022	26/07/2022
							F1905	Device Revision or Replacement	A0908	Incorrect, Inadequate or Imprecise Result or Readings	26/07/2022	26/07/2022
7553	2022	All other reportable incidents	17882	Physiologic-monitoring defibrillation system	CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A07	Electrical /Electronic Property Problem		
							F27	Problem Identified During Non-Clinical Procedure	A07	Electrical /Electronic Property Problem		
7554	2022	All other reportable incidents	47270	Cardiac resynchronization therapy implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A23	Use of Device Problem		
							F26	No Health Consequences or Impact	A23	Use of Device Problem		
7555	2022	All other reportable incidents	47270	Cardiac resynchronization therapy implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A072201	High Impedance		
							F26	No Health Consequences or Impact	A072201	High Impedance		
7556	2022	All other reportable incidents	35852	Single-chamber implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E2104	Electric Shock	A070909	Oversensing		
									A071302	Inappropriate/Inadequate Shock/Stimulation		
							E2403	No Clinical Signs, Symptoms or Conditions	A070909	Oversensing		

									A071302	Inappropriate/Inadequate Shock/Stimulation		
							F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A070909	Oversensing		
									A071302	Inappropriate/Inadequate Shock/Stimulation		
7557	2022	All other reportable incidents	35852	Single-chamber implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E2104	Electric Shock	A070909	Oversensing		
									A071302	Inappropriate/Inadequate Shock/Stimulation		
							E2403	No Clinical Signs, Symptoms or Conditions	A070909	Oversensing		
									A071302	Inappropriate/Inadequate Shock/Stimulation		
							F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A070909	Oversensing		
									A071302	Inappropriate/Inadequate Shock/Stimulation		
7558	2022	All other reportable incidents	35223	Endocardial/interventricular septal pacing lead	CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070901	Decreased Sensitivity		
									A070909	Oversensing		
							F12	Serious Injury/ Illness/ Impairment	A070901	Decreased Sensitivity		
									A070909	Oversensing		
							F23	Unexpected Medical Intervention	A070901	Decreased Sensitivity		
									A070909	Oversensing		
7559	2022	All other reportable incidents	17882	Physiologic-monitoring defibrillation system	CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A1603	Failure of Device to Self-Test		
							F27	Problem Identified During Non-Clinical Procedure	A1603	Failure of Device to Self-Test		
7560	2022	All other reportable incidents	35223	Endocardial/interventricular septal pacing lead	CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A072201	High Impedance	11/01/2022	11/01/2022
							F26	No Health Consequences or Impact	A072201	High Impedance	11/01/2022	11/01/2022
7561	2022	All other reportable incidents	47263	Cardiac resynchronizatlon therapy implantable pacemaker	CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	15/08/2022	15/08/2022
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	15/08/2022	15/08/2022
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	15/08/2022	15/08/2022
					CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	15/08/2022	15/08/2022
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	15/08/2022	15/08/2022
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	15/08/2022	15/08/2022
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	15/08/2022	15/08/2022
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	15/08/2022	15/08/2022
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	15/08/2022	15/08/2022
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	15/08/2022	15/08/2022
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	15/08/2022	15/08/2022
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	15/08/2022	15/08/2022
7562	2022	All other reportable incidents	35852	Single-chamber implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E2104	Electric Shock	A070909	Oversensing		
									A071302	Inappropriate/Inadequate Shock/Stimulation		
									A1301	Failure to Read Input Signal		
							E2403	No Clinical Signs, Symptoms or Conditions	A070909	Oversensing		
									A071302	Inappropriate/Inadequate Shock/Stimulation		
									A1301	Failure to Read Input Signal		
							F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A070909	Oversensing		
									A071302	Inappropriate/Inadequate Shock/Stimulation		
									A1301	Failure to Read Input Signal		
7563	2022	All other reportable incidents	47267	Single-chamber implantable pacemaker, rate-responsive	CT141	Electrodes/Leads and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	13/10/2022	13/10/2022
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	13/10/2022	13/10/2022
							F1903	Device Explantation	A070504	Premature Discharge of Battery	13/10/2022	13/10/2022
					CT145	Cardiovascular devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	13/10/2022	13/10/2022
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	13/10/2022	13/10/2022
							F1903	Device Explantation	A070504	Premature Discharge of Battery	13/10/2022	13/10/2022
					CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	13/10/2022	13/10/2022
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	13/10/2022	13/10/2022
							F1903	Device Explantation	A070504	Premature Discharge of Battery	13/10/2022	13/10/2022
					CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	13/10/2022	13/10/2022
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	13/10/2022	13/10/2022
							F1903	Device Explantation	A070504	Premature Discharge of Battery	13/10/2022	13/10/2022
					CT233	Surgical	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	13/10/2022	13/10/2022
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	13/10/2022	13/10/2022
							F1903	Device Explantation	A070504	Premature Discharge of Battery	13/10/2022	13/10/2022
					CT2406	Implantable	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	13/10/2022	13/10/2022
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	13/10/2022	13/10/2022
							F1903	Device Explantation	A070504	Premature Discharge of Battery	13/10/2022	13/10/2022
					CT2407	Active implantable	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	13/10/2022	13/10/2022
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	13/10/2022	13/10/2022
							F1903	Device Explantation	A070504	Premature Discharge of Battery	13/10/2022	13/10/2022
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	13/10/2022	13/10/2022
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	13/10/2022	13/10/2022
							F1903	Device Explantation	A070504	Premature Discharge of Battery	13/10/2022	13/10/2022
					CT2727	Active implant devices/ancillaries	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	13/10/2022	13/10/2022
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	13/10/2022	13/10/2022
							F1903	Device Explantation	A070504	Premature Discharge of Battery	13/10/2022	13/10/2022
					CT2750	Implantable cardiac electrical stimulators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	13/10/2022	13/10/2022
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	13/10/2022	13/10/2022
							F1903	Device Explantation	A070504	Premature Discharge of Battery	13/10/2022	13/10/2022
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	13/10/2022	13/10/2022
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	13/10/2022	13/10/2022
							F1903	Device Explantation	A070504	Premature Discharge of Battery	13/10/2022	13/10/2022
					CT2985	Active	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	13/10/2022	13/10/2022
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	13/10/2022	13/10/2022
							F1903	Device Explantation	A070504	Premature Discharge of Battery	13/10/2022	13/10/2022
					CT3001	Implantable metal	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	13/10/2022	13/10/2022
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	13/10/2022	13/10/2022
							F1903	Device Explantation	A070504	Premature Discharge of Battery	13/10/2022	13/10/2022
					CT3002	Implantable synthetic polymer	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	13/10/2022	13/10/2022
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	13/10/2022	13/10/2022
							F1903	Device Explantation	A070504	Premature Discharge of Battery	13/10/2022	13/10/2022
					CT3006	Implantable inorganic material	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	13/10/2022	13/10/2022
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	13/10/2022	13/10/2022
							F1903	Device Explantation	A070504	Premature Discharge of Battery	13/10/2022	13/10/2022
					CT314	Electrically powered	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	13/10/2022	13/10/2022
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	13/10/2022	13/10/2022
							F1903	Device Explantation	A070504	Premature Discharge of Battery	13/10/2022	13/10/2022

					CT321	Long-term surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	13/10/2022	13/10/2022
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	13/10/2022	13/10/2022
							F1903	Device Explantation	A070504	Premature Discharge of Battery	13/10/2022	13/10/2022
					CT449	Generators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	13/10/2022	13/10/2022
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	13/10/2022	13/10/2022
							F1903	Device Explantation	A070504	Premature Discharge of Battery	13/10/2022	13/10/2022
					CT981	Single use	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	13/10/2022	13/10/2022
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	13/10/2022	13/10/2022
							F1903	Device Explantation	A070504	Premature Discharge of Battery	13/10/2022	13/10/2022
					CT983	Surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	13/10/2022	13/10/2022
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	13/10/2022	13/10/2022
							F1903	Device Explantation	A070504	Premature Discharge of Battery	13/10/2022	13/10/2022
7564	2022	All other reportable incidents	35852	Single-chamber implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F24	Insufficient Health Impact Information	A070504	Premature Discharge of Battery		
7565	2022	All other reportable incidents	36241	Pacemaker/defibrillator lead	CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A0203	Defective Device	15/09/2022	15/09/2022
							F08	Hospitalization or Prolonged Hospitalization	A0203	Defective Device	15/09/2022	15/09/2022
							F1905	Device Revision or Replacement	A0203	Defective Device	15/09/2022	15/09/2022
7566	2022	All other reportable incidents	37265	Dual-chamber implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	29/08/2022	29/08/2022
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	29/08/2022	29/08/2022
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	29/08/2022	29/08/2022
7567	2022	All other reportable incidents	35223	Endocardial/interventricular septal pacing lead	CT1684	Cardiac pacemakers and associated devices	E011903	Syncope/Fainting	A051201	Device Dislodged or Dislocated	13/09/2022	13/09/2022
							E0601	Arrhythmia	A051201	Device Dislodged or Dislocated	13/09/2022	13/09/2022
							E2403	No Clinical Signs, Symptoms or Conditions	A051201	Device Dislodged or Dislocated	13/09/2022	13/09/2022
							F1901	Additional Surgery	A051201	Device Dislodged or Dislocated	13/09/2022	13/09/2022
7568	2022	All other reportable incidents	17882	Physiologic-monitoring defibrillation system	CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem		
							F27	Problem Identified During Non-Clinical Procedure	A0705	Battery Problem		
7569	2022	Unanticipated serious deterioration in state of health	47265	Dual-chamber implantable pacemaker, rate-responsive	CT141	Electrodes/Leads and associated devices	E2401	Insufficient Information	A071205	Pacemaker Found in Back-Up Mode	25/08/2022	
							F24	Insufficient Health Impact Information	A071205	Pacemaker Found in Back-Up Mode	25/08/2022	
					CT145	Cardiovascular devices	E2401	Insufficient Information	A071205	Pacemaker Found in Back-Up Mode	25/08/2022	
							F24	Insufficient Health Impact Information	A071205	Pacemaker Found in Back-Up Mode	25/08/2022	
					CT1684	Cardiac pacemakers and associated devices	E2401	Insufficient Information	A071205	Pacemaker Found in Back-Up Mode	25/08/2022	
							F24	Insufficient Health Impact Information	A071205	Pacemaker Found in Back-Up Mode	25/08/2022	
					CT1685	Cardiac pacemakers	E2401	Insufficient Information	A071205	Pacemaker Found in Back-Up Mode	25/08/2022	
							F24	Insufficient Health Impact Information	A071205	Pacemaker Found in Back-Up Mode	25/08/2022	
					CT233	Surgical	E2401	Insufficient Information	A071205	Pacemaker Found in Back-Up Mode	25/08/2022	
							F24	Insufficient Health Impact Information	A071205	Pacemaker Found in Back-Up Mode	25/08/2022	
					CT2406	Implantable	E2401	Insufficient Information	A071205	Pacemaker Found in Back-Up Mode	25/08/2022	
							F24	Insufficient Health Impact Information	A071205	Pacemaker Found in Back-Up Mode	25/08/2022	
					CT2407	Active implantable	E2401	Insufficient Information	A071205	Pacemaker Found in Back-Up Mode	25/08/2022	
							F24	Insufficient Health Impact Information	A071205	Pacemaker Found in Back-Up Mode	25/08/2022	
					CT257	Implantable cardiac pacemakers	E2401	Insufficient Information	A071205	Pacemaker Found in Back-Up Mode	25/08/2022	
							F24	Insufficient Health Impact Information	A071205	Pacemaker Found in Back-Up Mode	25/08/2022	
					CT2727	Active implant devices/ancillaries	E2401	Insufficient Information	A071205	Pacemaker Found in Back-Up Mode	25/08/2022	
							F24	Insufficient Health Impact Information	A071205	Pacemaker Found in Back-Up Mode	25/08/2022	
					CT2750	Implantable cardiac electrical stimulators and associated devices	E2401	Insufficient Information	A071205	Pacemaker Found in Back-Up Mode	25/08/2022	
							F24	Insufficient Health Impact Information	A071205	Pacemaker Found in Back-Up Mode	25/08/2022	
					CT2866	Implantable cardiac electrical stimulators	E2401	Insufficient Information	A071205	Pacemaker Found in Back-Up Mode	25/08/2022	
							F24	Insufficient Health Impact Information	A071205	Pacemaker Found in Back-Up Mode	25/08/2022	
					CT2985	Active	E2401	Insufficient Information	A071205	Pacemaker Found in Back-Up Mode	25/08/2022	
							F24	Insufficient Health Impact Information	A071205	Pacemaker Found in Back-Up Mode	25/08/2022	
					CT3001	Implantable metal	E2401	Insufficient Information	A071205	Pacemaker Found in Back-Up Mode	25/08/2022	
							F24	Insufficient Health Impact Information	A071205	Pacemaker Found in Back-Up Mode	25/08/2022	
					CT3002	Implantable synthetic polymer	E2401	Insufficient Information	A071205	Pacemaker Found in Back-Up Mode	25/08/2022	
							F24	Insufficient Health Impact Information	A071205	Pacemaker Found in Back-Up Mode	25/08/2022	
					CT3006	Implantable inorganic material	E2401	Insufficient Information	A071205	Pacemaker Found in Back-Up Mode	25/08/2022	
							F24	Insufficient Health Impact Information	A071205	Pacemaker Found in Back-Up Mode	25/08/2022	
					CT314	Electrically powered	E2401	Insufficient Information	A071205	Pacemaker Found in Back-Up Mode	25/08/2022	
							F24	Insufficient Health Impact Information	A071205	Pacemaker Found in Back-Up Mode	25/08/2022	
					CT321	Long-term surgical invasive	E2401	Insufficient Information	A071205	Pacemaker Found in Back-Up Mode	25/08/2022	
							F24	Insufficient Health Impact Information	A071205	Pacemaker Found in Back-Up Mode	25/08/2022	
					CT449	Generators and associated devices	E2401	Insufficient Information	A071205	Pacemaker Found in Back-Up Mode	25/08/2022	
							F24	Insufficient Health Impact Information	A071205	Pacemaker Found in Back-Up Mode	25/08/2022	
					CT981	Single use	E2401	Insufficient Information	A071205	Pacemaker Found in Back-Up Mode	25/08/2022	
							F24	Insufficient Health Impact Information	A071205	Pacemaker Found in Back-Up Mode	25/08/2022	
					CT983	Surgical invasive	E2401	Insufficient Information	A071205	Pacemaker Found in Back-Up Mode	25/08/2022	
							F24	Insufficient Health Impact Information	A071205	Pacemaker Found in Back-Up Mode	25/08/2022	
7570	2022	All other reportable incidents	47265	Dual-chamber implantable pacemaker, rate-responsive	CT141	Electrodes/Leads and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	22/02/2022	22/02/2022
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	22/02/2022	22/02/2022
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	22/02/2022	22/02/2022
					CT145	Cardiovascular devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	22/02/2022	22/02/2022
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	22/02/2022	22/02/2022
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	22/02/2022	22/02/2022
					CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	22/02/2022	22/02/2022
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	22/02/2022	22/02/2022
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	22/02/2022	22/02/2022
					CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	22/02/2022	22/02/2022
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	22/02/2022	22/02/2022
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	22/02/2022	22/02/2022
					CT233	Surgical	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	22/02/2022	22/02/2022
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	22/02/2022	22/02/2022
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	22/02/2022	22/02/2022
					CT2406	Implantable	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	22/02/2022	22/02/2022
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	22/02/2022	22/02/2022
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	22/02/2022	22/02/2022
					CT2407	Active implantable	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	22/02/2022	22/02/2022
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	22/02/2022	22/02/2022
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	22/02/2022	22/02/2022
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	22/02/2022	22/02/2022

							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	22/02/2022	22/02/2022
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	22/02/2022	22/02/2022
					CT2727	Active implant devices/ancillaries	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	22/02/2022	22/02/2022
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	22/02/2022	22/02/2022
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	22/02/2022	22/02/2022
					CT2750	Implantable cardiac electrical stimulators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	22/02/2022	22/02/2022
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	22/02/2022	22/02/2022
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	22/02/2022	22/02/2022
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	22/02/2022	22/02/2022
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	22/02/2022	22/02/2022
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	22/02/2022	22/02/2022
					CT2985	Active	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	22/02/2022	22/02/2022
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	22/02/2022	22/02/2022
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	22/02/2022	22/02/2022
					CT3001	Implantable metal	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	22/02/2022	22/02/2022
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	22/02/2022	22/02/2022
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	22/02/2022	22/02/2022
					CT3002	Implantable synthetic polymer	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	22/02/2022	22/02/2022
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	22/02/2022	22/02/2022
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	22/02/2022	22/02/2022
					CT3006	Implantable inorganic material	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	22/02/2022	22/02/2022
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	22/02/2022	22/02/2022
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	22/02/2022	22/02/2022
					CT314	Electrically powered	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	22/02/2022	22/02/2022
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	22/02/2022	22/02/2022
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	22/02/2022	22/02/2022
					CT321	Long-term surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	22/02/2022	22/02/2022
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	22/02/2022	22/02/2022
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	22/02/2022	22/02/2022
					CT449	Generators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	22/02/2022	22/02/2022
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	22/02/2022	22/02/2022
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	22/02/2022	22/02/2022
					CT981	Single use	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	22/02/2022	22/02/2022
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	22/02/2022	22/02/2022
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	22/02/2022	22/02/2022
					CT983	Surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	22/02/2022	22/02/2022
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	22/02/2022	22/02/2022
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	22/02/2022	22/02/2022
7571	2022	All other reportable incidents	47270	Cardiac resynchronization therapy implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
7572	2022	All other reportable incidents	37265	Dual-chamber implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E0623	Unspecified Heart Problem	A071302	Inappropriate/Inadequate Shock/Stimulation		
							E2104	Electric Shock	A071302	Inappropriate/Inadequate Shock/Stimulation		
							E2403	No Clinical Signs, Symptoms or Conditions	A071302	Inappropriate/Inadequate Shock/Stimulation		
							F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A071302	Inappropriate/Inadequate Shock/Stimulation		
7573	2022	All other reportable incidents	47265	Dual-chamber implantable pacemaker, rate-responsive	CT141	Electrodes/Leads and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	22/08/2022	22/08/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	22/08/2022	22/08/2022
					CT145	Cardiovascular devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	22/08/2022	22/08/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	22/08/2022	22/08/2022
					CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	22/08/2022	22/08/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	22/08/2022	22/08/2022
					CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	22/08/2022	22/08/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	22/08/2022	22/08/2022
					CT233	Surgical	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	22/08/2022	22/08/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	22/08/2022	22/08/2022
					CT2406	Implantable	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	22/08/2022	22/08/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	22/08/2022	22/08/2022
					CT2407	Active implantable	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	22/08/2022	22/08/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	22/08/2022	22/08/2022
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	22/08/2022	22/08/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	22/08/2022	22/08/2022
					CT2727	Active implant devices/ancillaries	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	22/08/2022	22/08/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	22/08/2022	22/08/2022
					CT2750	Implantable cardiac electrical stimulators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	22/08/2022	22/08/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	22/08/2022	22/08/2022
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	22/08/2022	22/08/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	22/08/2022	22/08/2022
					CT2985	Active	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	22/08/2022	22/08/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	22/08/2022	22/08/2022
					CT3001	Implantable metal	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	22/08/2022	22/08/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	22/08/2022	22/08/2022
					CT3002	Implantable synthetic polymer	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	22/08/2022	22/08/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	22/08/2022	22/08/2022
					CT3006	Implantable inorganic material	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	22/08/2022	22/08/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	22/08/2022	22/08/2022
					CT314	Electrically powered	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	22/08/2022	22/08/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	22/08/2022	22/08/2022
					CT321	Long-term surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	22/08/2022	22/08/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	22/08/2022	22/08/2022
					CT449	Generators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	22/08/2022	22/08/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	22/08/2022	22/08/2022
					CT981	Single use	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	22/08/2022	22/08/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	22/08/2022	22/08/2022
					CT983	Surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	22/08/2022	22/08/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	22/08/2022	22/08/2022
7574	2022	All other reportable incidents	47263	Cardiac resynchronization therapy implantable pacemaker	CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	18/08/2022	18/08/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	18/08/2022	18/08/2022
					CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	18/08/2022	18/08/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	18/08/2022	18/08/2022

					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	18/08/2022	18/08/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	18/08/2022	18/08/2022
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	18/08/2022	18/08/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	18/08/2022	18/08/2022
7575	2022	All other reportable incidents	47265	Dual-chamber implantable pacemaker, rate-responsive	CT141	Electrodes/Leads and associated devices	E0601	Arrhythmia	A070504	Premature Discharge of Battery	23/08/2022	23/08/2022
									A071102	Failure to Interrogate	23/08/2022	23/08/2022
									A071204	No Pacing	23/08/2022	23/08/2022
									A1305	Wireless Communication Problem	23/08/2022	23/08/2022
							E2311	Discomfort	A070504	Premature Discharge of Battery	23/08/2022	23/08/2022
									A071102	Failure to Interrogate	23/08/2022	23/08/2022
									A071204	No Pacing	23/08/2022	23/08/2022
									A1305	Wireless Communication Problem	23/08/2022	23/08/2022
							E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	23/08/2022	23/08/2022
									A071102	Failure to Interrogate	23/08/2022	23/08/2022
									A071204	No Pacing	23/08/2022	23/08/2022
									A1305	Wireless Communication Problem	23/08/2022	23/08/2022
							F1901	Additional Surgery	A070504	Premature Discharge of Battery	23/08/2022	23/08/2022
									A071102	Failure to Interrogate	23/08/2022	23/08/2022
									A071204	No Pacing	23/08/2022	23/08/2022
									A1305	Wireless Communication Problem	23/08/2022	23/08/2022
					CT145	Cardiovascular devices	E0601	Arrhythmia	A070504	Premature Discharge of Battery	23/08/2022	23/08/2022
									A071102	Failure to Interrogate	23/08/2022	23/08/2022
									A071204	No Pacing	23/08/2022	23/08/2022
									A1305	Wireless Communication Problem	23/08/2022	23/08/2022
							E2311	Discomfort	A070504	Premature Discharge of Battery	23/08/2022	23/08/2022
									A071102	Failure to Interrogate	23/08/2022	23/08/2022
									A071204	No Pacing	23/08/2022	23/08/2022
									A1305	Wireless Communication Problem	23/08/2022	23/08/2022
							E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	23/08/2022	23/08/2022
									A071102	Failure to Interrogate	23/08/2022	23/08/2022
									A071204	No Pacing	23/08/2022	23/08/2022
									A1305	Wireless Communication Problem	23/08/2022	23/08/2022
							F1901	Additional Surgery	A070504	Premature Discharge of Battery	23/08/2022	23/08/2022
									A071102	Failure to Interrogate	23/08/2022	23/08/2022
									A071204	No Pacing	23/08/2022	23/08/2022
									A1305	Wireless Communication Problem	23/08/2022	23/08/2022
					CT1684	Cardiac pacemakers and associated devices	E0601	Arrhythmia	A070504	Premature Discharge of Battery	23/08/2022	23/08/2022
									A071102	Failure to Interrogate	23/08/2022	23/08/2022
									A071204	No Pacing	23/08/2022	23/08/2022
									A1305	Wireless Communication Problem	23/08/2022	23/08/2022
							E2311	Discomfort	A070504	Premature Discharge of Battery	23/08/2022	23/08/2022
									A071102	Failure to Interrogate	23/08/2022	23/08/2022
									A071204	No Pacing	23/08/2022	23/08/2022
									A1305	Wireless Communication Problem	23/08/2022	23/08/2022
							E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	23/08/2022	23/08/2022
									A071102	Failure to Interrogate	23/08/2022	23/08/2022
									A071204	No Pacing	23/08/2022	23/08/2022
									A1305	Wireless Communication Problem	23/08/2022	23/08/2022
							F1901	Additional Surgery	A070504	Premature Discharge of Battery	23/08/2022	23/08/2022
									A071102	Failure to Interrogate	23/08/2022	23/08/2022
									A071204	No Pacing	23/08/2022	23/08/2022
									A1305	Wireless Communication Problem	23/08/2022	23/08/2022
					CT1685	Cardiac pacemakers	E0601	Arrhythmia	A070504	Premature Discharge of Battery	23/08/2022	23/08/2022
									A071102	Failure to Interrogate	23/08/2022	23/08/2022
									A071204	No Pacing	23/08/2022	23/08/2022
									A1305	Wireless Communication Problem	23/08/2022	23/08/2022
							E2311	Discomfort	A070504	Premature Discharge of Battery	23/08/2022	23/08/2022
									A071102	Failure to Interrogate	23/08/2022	23/08/2022
									A071204	No Pacing	23/08/2022	23/08/2022
									A1305	Wireless Communication Problem	23/08/2022	23/08/2022
							E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	23/08/2022	23/08/2022
									A071102	Failure to Interrogate	23/08/2022	23/08/2022
									A071204	No Pacing	23/08/2022	23/08/2022
									A1305	Wireless Communication Problem	23/08/2022	23/08/2022
							F1901	Additional Surgery	A070504	Premature Discharge of Battery	23/08/2022	23/08/2022
									A071102	Failure to Interrogate	23/08/2022	23/08/2022
									A071204	No Pacing	23/08/2022	23/08/2022
									A1305	Wireless Communication Problem	23/08/2022	23/08/2022
					CT233	Surgical	E0601	Arrhythmia	A070504	Premature Discharge of Battery	23/08/2022	23/08/2022
									A071102	Failure to Interrogate	23/08/2022	23/08/2022
									A071204	No Pacing	23/08/2022	23/08/2022
									A1305	Wireless Communication Problem	23/08/2022	23/08/2022
							E2311	Discomfort	A070504	Premature Discharge of Battery	23/08/2022	23/08/2022
									A071102	Failure to Interrogate	23/08/2022	23/08/2022
									A071204	No Pacing	23/08/2022	23/08/2022
									A1305	Wireless Communication Problem	23/08/2022	23/08/2022
							E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	23/08/2022	23/08/2022
									A071102	Failure to Interrogate	23/08/2022	23/08/2022
									A071204	No Pacing	23/08/2022	23/08/2022
									A1305	Wireless Communication Problem	23/08/2022	23/08/2022
							F1901	Additional Surgery	A070504	Premature Discharge of Battery	23/08/2022	23/08/2022
									A071102	Failure to Interrogate	23/08/2022	23/08/2022
									A071204	No Pacing	23/08/2022	23/08/2022
									A1305	Wireless Communication Problem	23/08/2022	23/08/2022
					CT2406	Implantable	E0601	Arrhythmia	A070504	Premature Discharge of Battery	23/08/2022	23/08/2022
									A071102	Failure to Interrogate	23/08/2022	23/08/2022
									A071204	No Pacing	23/08/2022	23/08/2022
									A1305	Wireless Communication Problem	23/08/2022	23/08/2022

							E2311	Discomfort	A070504	Premature Discharge of Battery	23/08/2022	23/08/2022
									A071102	Failure to Interrogate	23/08/2022	23/08/2022
									A071204	No Pacing	23/08/2022	23/08/2022
									A1305	Wireless Communication Problem	23/08/2022	23/08/2022
							E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	23/08/2022	23/08/2022
									A071102	Failure to Interrogate	23/08/2022	23/08/2022
									A071204	No Pacing	23/08/2022	23/08/2022
									A1305	Wireless Communication Problem	23/08/2022	23/08/2022
							F1901	Additional Surgery	A070504	Premature Discharge of Battery	23/08/2022	23/08/2022
									A071102	Failure to Interrogate	23/08/2022	23/08/2022
									A071204	No Pacing	23/08/2022	23/08/2022
									A1305	Wireless Communication Problem	23/08/2022	23/08/2022
						CT2407	Active implantable	E0601	Arrhythmia	A070504	Premature Discharge of Battery	23/08/2022	23/08/2022
										A071102	Failure to Interrogate	23/08/2022	23/08/2022
										A071204	No Pacing	23/08/2022	23/08/2022
										A1305	Wireless Communication Problem	23/08/2022	23/08/2022
								E2311	Discomfort	A070504	Premature Discharge of Battery	23/08/2022	23/08/2022
										A071102	Failure to Interrogate	23/08/2022	23/08/2022
										A071204	No Pacing	23/08/2022	23/08/2022
										A1305	Wireless Communication Problem	23/08/2022	23/08/2022
								E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	23/08/2022	23/08/2022
										A071102	Failure to Interrogate	23/08/2022	23/08/2022
										A071204	No Pacing	23/08/2022	23/08/2022
										A1305	Wireless Communication Problem	23/08/2022	23/08/2022
								F1901	Additional Surgery	A070504	Premature Discharge of Battery	23/08/2022	23/08/2022
										A071102	Failure to Interrogate	23/08/2022	23/08/2022
										A071204	No Pacing	23/08/2022	23/08/2022
										A1305	Wireless Communication Problem	23/08/2022	23/08/2022
						CT257	Implantable cardiac pacemakers	E0601	Arrhythmia	A070504	Premature Discharge of Battery	23/08/2022	23/08/2022
										A071102	Failure to Interrogate	23/08/2022	23/08/2022
										A071204	No Pacing	23/08/2022	23/08/2022
										A1305	Wireless Communication Problem	23/08/2022	23/08/2022
								E2311	Discomfort	A070504	Premature Discharge of Battery	23/08/2022	23/08/2022
										A071102	Failure to Interrogate	23/08/2022	23/08/2022
										A071204	No Pacing	23/08/2022	23/08/2022
										A1305	Wireless Communication Problem	23/08/2022	23/08/2022
								E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	23/08/2022	23/08/2022
										A071102	Failure to Interrogate	23/08/2022	23/08/2022
										A071204	No Pacing	23/08/2022	23/08/2022
										A1305	Wireless Communication Problem	23/08/2022	23/08/2022
								F1901	Additional Surgery	A070504	Premature Discharge of Battery	23/08/2022	23/08/2022
										A071102	Failure to Interrogate	23/08/2022	23/08/2022
										A071204	No Pacing	23/08/2022	23/08/2022
										A1305	Wireless Communication Problem	23/08/2022	23/08/2022
						CT2727	Active implant devices/ancillaries	E0601	Arrhythmia	A070504	Premature Discharge of Battery	23/08/2022	23/08/2022
										A071102	Failure to Interrogate	23/08/2022	23/08/2022
										A071204	No Pacing	23/08/2022	23/08/2022
										A1305	Wireless Communication Problem	23/08/2022	23/08/2022
								E2311	Discomfort	A070504	Premature Discharge of Battery	23/08/2022	23/08/2022
										A071102	Failure to Interrogate	23/08/2022	23/08/2022
										A071204	No Pacing	23/08/2022	23/08/2022
										A1305	Wireless Communication Problem	23/08/2022	23/08/2022
								E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	23/08/2022	23/08/2022
										A071102	Failure to Interrogate	23/08/2022	23/08/2022
										A071204	No Pacing	23/08/2022	23/08/2022
										A1305	Wireless Communication Problem	23/08/2022	23/08/2022
								F1901	Additional Surgery	A070504	Premature Discharge of Battery	23/08/2022	23/08/2022
										A071102	Failure to Interrogate	23/08/2022	23/08/2022
										A071204	No Pacing	23/08/2022	23/08/2022
										A1305	Wireless Communication Problem	23/08/2022	23/08/2022
						CT2750	Implantable cardiac electrical stimulators and associated devices	E0601	Arrhythmia	A070504	Premature Discharge of Battery	23/08/2022	23/08/2022
										A071102	Failure to Interrogate	23/08/2022	23/08/2022
										A071204	No Pacing	23/08/2022	23/08/2022
										A1305	Wireless Communication Problem	23/08/2022	23/08/2022
								E2311	Discomfort	A070504	Premature Discharge of Battery	23/08/2022	23/08/2022
										A071102	Failure to Interrogate	23/08/2022	23/08/2022
										A071204	No Pacing	23/08/2022	23/08/2022
										A1305	Wireless Communication Problem	23/08/2022	23/08/2022
								E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	23/08/2022	23/08/2022
										A071102	Failure to Interrogate	23/08/2022	23/08/2022
										A071204	No Pacing	23/08/2022	23/08/2022
										A1305	Wireless Communication Problem	23/08/2022	23/08/2022
								F1901	Additional Surgery	A070504	Premature Discharge of Battery	23/08/2022	23/08/2022
										A071102	Failure to Interrogate	23/08/2022	23/08/2022
										A071204	No Pacing	23/08/2022	23/08/2022
										A1305	Wireless Communication Problem	23/08/2022	23/08/2022
						CT2866	Implantable cardiac electrical stimulators	E0601	Arrhythmia	A070504	Premature Discharge of Battery	23/08/2022	23/08/2022
										A071102	Failure to Interrogate	23/08/2022	23/08/2022
										A071204	No Pacing	23/08/2022	23/08/2022
										A1305	Wireless Communication Problem	23/08/2022	23/08/2022
								E2311	Discomfort	A070504	Premature Discharge of Battery	23/08/2022	23/08/2022
										A071102	Failure to Interrogate	23/08/2022	23/08/2022
										A071204	No Pacing	23/08/2022	23/08/2022
										A1305	Wireless Communication Problem	23/08/2022	23/08/2022
								E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	23/08/2022	23/08/2022
										A071102	Failure to Interrogate	23/08/2022	23/08/2022
										A071204	No Pacing	23/08/2022	23/08/2022
										A1305	Wireless Communication Problem	23/08/2022	23/08/2022

							F1901	Additional Surgery	A070504	Premature Discharge of Battery	23/08/2022	23/08/2022
									A071102	Failure to Interrogate	23/08/2022	23/08/2022
									A071204	No Pacing	23/08/2022	23/08/2022
									A1305	Wireless Communication Problem	23/08/2022	23/08/2022
					CT2985	Active	E0601	Arrhythmia	A070504	Premature Discharge of Battery	23/08/2022	23/08/2022
									A071102	Failure to Interrogate	23/08/2022	23/08/2022
									A071204	No Pacing	23/08/2022	23/08/2022
									A1305	Wireless Communication Problem	23/08/2022	23/08/2022
							E2311	Discomfort	A070504	Premature Discharge of Battery	23/08/2022	23/08/2022
									A071102	Failure to Interrogate	23/08/2022	23/08/2022
									A071204	No Pacing	23/08/2022	23/08/2022
									A1305	Wireless Communication Problem	23/08/2022	23/08/2022
							E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	23/08/2022	23/08/2022
									A071102	Failure to Interrogate	23/08/2022	23/08/2022
									A071204	No Pacing	23/08/2022	23/08/2022
									A1305	Wireless Communication Problem	23/08/2022	23/08/2022
							F1901	Additional Surgery	A070504	Premature Discharge of Battery	23/08/2022	23/08/2022
									A071102	Failure to Interrogate	23/08/2022	23/08/2022
									A071204	No Pacing	23/08/2022	23/08/2022
									A1305	Wireless Communication Problem	23/08/2022	23/08/2022
					CT3001	Implantable metal	E0601	Arrhythmia	A070504	Premature Discharge of Battery	23/08/2022	23/08/2022
									A071102	Failure to Interrogate	23/08/2022	23/08/2022
									A071204	No Pacing	23/08/2022	23/08/2022
									A1305	Wireless Communication Problem	23/08/2022	23/08/2022
							E2311	Discomfort	A070504	Premature Discharge of Battery	23/08/2022	23/08/2022
									A071102	Failure to Interrogate	23/08/2022	23/08/2022
									A071204	No Pacing	23/08/2022	23/08/2022
									A1305	Wireless Communication Problem	23/08/2022	23/08/2022
							E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	23/08/2022	23/08/2022
									A071102	Failure to Interrogate	23/08/2022	23/08/2022
									A071204	No Pacing	23/08/2022	23/08/2022
									A1305	Wireless Communication Problem	23/08/2022	23/08/2022
							F1901	Additional Surgery	A070504	Premature Discharge of Battery	23/08/2022	23/08/2022
									A071102	Failure to Interrogate	23/08/2022	23/08/2022
									A071204	No Pacing	23/08/2022	23/08/2022
									A1305	Wireless Communication Problem	23/08/2022	23/08/2022
					CT3002	Implantable synthetic polymer	E0601	Arrhythmia	A070504	Premature Discharge of Battery	23/08/2022	23/08/2022
									A071102	Failure to Interrogate	23/08/2022	23/08/2022
									A071204	No Pacing	23/08/2022	23/08/2022
									A1305	Wireless Communication Problem	23/08/2022	23/08/2022
							E2311	Discomfort	A070504	Premature Discharge of Battery	23/08/2022	23/08/2022
									A071102	Failure to Interrogate	23/08/2022	23/08/2022
									A071204	No Pacing	23/08/2022	23/08/2022
									A1305	Wireless Communication Problem	23/08/2022	23/08/2022
							E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	23/08/2022	23/08/2022
									A071102	Failure to Interrogate	23/08/2022	23/08/2022
									A071204	No Pacing	23/08/2022	23/08/2022
									A1305	Wireless Communication Problem	23/08/2022	23/08/2022
							F1901	Additional Surgery	A070504	Premature Discharge of Battery	23/08/2022	23/08/2022
									A071102	Failure to Interrogate	23/08/2022	23/08/2022
									A071204	No Pacing	23/08/2022	23/08/2022
									A1305	Wireless Communication Problem	23/08/2022	23/08/2022
					CT3006	Implantable inorganic material	E0601	Arrhythmia	A070504	Premature Discharge of Battery	23/08/2022	23/08/2022
									A071102	Failure to Interrogate	23/08/2022	23/08/2022
									A071204	No Pacing	23/08/2022	23/08/2022
									A1305	Wireless Communication Problem	23/08/2022	23/08/2022
							E2311	Discomfort	A070504	Premature Discharge of Battery	23/08/2022	23/08/2022
									A071102	Failure to Interrogate	23/08/2022	23/08/2022
									A071204	No Pacing	23/08/2022	23/08/2022
									A1305	Wireless Communication Problem	23/08/2022	23/08/2022
							E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	23/08/2022	23/08/2022
									A071102	Failure to Interrogate	23/08/2022	23/08/2022
									A071204	No Pacing	23/08/2022	23/08/2022
									A1305	Wireless Communication Problem	23/08/2022	23/08/2022
							F1901	Additional Surgery	A070504	Premature Discharge of Battery	23/08/2022	23/08/2022
									A071102	Failure to Interrogate	23/08/2022	23/08/2022
									A071204	No Pacing	23/08/2022	23/08/2022
									A1305	Wireless Communication Problem	23/08/2022	23/08/2022
					CT314	Electrically powered	E0601	Arrhythmia	A070504	Premature Discharge of Battery	23/08/2022	23/08/2022
									A071102	Failure to Interrogate	23/08/2022	23/08/2022
									A071204	No Pacing	23/08/2022	23/08/2022
									A1305	Wireless Communication Problem	23/08/2022	23/08/2022
							E2311	Discomfort	A070504	Premature Discharge of Battery	23/08/2022	23/08/2022
									A071102	Failure to Interrogate	23/08/2022	23/08/2022
									A071204	No Pacing	23/08/2022	23/08/2022
									A1305	Wireless Communication Problem	23/08/2022	23/08/2022
							E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	23/08/2022	23/08/2022
									A071102	Failure to Interrogate	23/08/2022	23/08/2022
									A071204	No Pacing	23/08/2022	23/08/2022
									A1305	Wireless Communication Problem	23/08/2022	23/08/2022
							F1901	Additional Surgery	A070504	Premature Discharge of Battery	23/08/2022	23/08/2022
									A071102	Failure to Interrogate	23/08/2022	23/08/2022
									A071204	No Pacing	23/08/2022	23/08/2022
									A1305	Wireless Communication Problem	23/08/2022	23/08/2022
					CT321	Long-term surgical invasive	E0601	Arrhythmia	A070504	Premature Discharge of Battery	23/08/2022	23/08/2022
									A071102	Failure to Interrogate	23/08/2022	23/08/2022
									A071204	No Pacing	23/08/2022	23/08/2022
									A1305	Wireless Communication Problem	23/08/2022	23/08/2022

							E2311	Discomfort	A070504	Premature Discharge of Battery	23/08/2022	23/08/2022
									A071102	Failure to Interrogate	23/08/2022	23/08/2022
									A071204	No Pacing	23/08/2022	23/08/2022
									A1305	Wireless Communication Problem	23/08/2022	23/08/2022
							E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	23/08/2022	23/08/2022
									A071102	Failure to Interrogate	23/08/2022	23/08/2022
									A071204	No Pacing	23/08/2022	23/08/2022
									A1305	Wireless Communication Problem	23/08/2022	23/08/2022
							F1901	Additional Surgery	A070504	Premature Discharge of Battery	23/08/2022	23/08/2022
									A071102	Failure to Interrogate	23/08/2022	23/08/2022
									A071204	No Pacing	23/08/2022	23/08/2022
									A1305	Wireless Communication Problem	23/08/2022	23/08/2022
					CT449	Generators and associated devices	E0601	Arrhythmia	A070504	Premature Discharge of Battery	23/08/2022	23/08/2022
									A071102	Failure to Interrogate	23/08/2022	23/08/2022
									A071204	No Pacing	23/08/2022	23/08/2022
									A1305	Wireless Communication Problem	23/08/2022	23/08/2022
							E2311	Discomfort	A070504	Premature Discharge of Battery	23/08/2022	23/08/2022
									A071102	Failure to Interrogate	23/08/2022	23/08/2022
									A071204	No Pacing	23/08/2022	23/08/2022
									A1305	Wireless Communication Problem	23/08/2022	23/08/2022
							E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	23/08/2022	23/08/2022
									A071102	Failure to Interrogate	23/08/2022	23/08/2022
									A071204	No Pacing	23/08/2022	23/08/2022
									A1305	Wireless Communication Problem	23/08/2022	23/08/2022
							F1901	Additional Surgery	A070504	Premature Discharge of Battery	23/08/2022	23/08/2022
									A071102	Failure to Interrogate	23/08/2022	23/08/2022
									A071204	No Pacing	23/08/2022	23/08/2022
									A1305	Wireless Communication Problem	23/08/2022	23/08/2022
					CT981	Single use	E0601	Arrhythmia	A070504	Premature Discharge of Battery	23/08/2022	23/08/2022
									A071102	Failure to Interrogate	23/08/2022	23/08/2022
									A071204	No Pacing	23/08/2022	23/08/2022
									A1305	Wireless Communication Problem	23/08/2022	23/08/2022
							E2311	Discomfort	A070504	Premature Discharge of Battery	23/08/2022	23/08/2022
									A071102	Failure to Interrogate	23/08/2022	23/08/2022
									A071204	No Pacing	23/08/2022	23/08/2022
									A1305	Wireless Communication Problem	23/08/2022	23/08/2022
							E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	23/08/2022	23/08/2022
									A071102	Failure to Interrogate	23/08/2022	23/08/2022
									A071204	No Pacing	23/08/2022	23/08/2022
									A1305	Wireless Communication Problem	23/08/2022	23/08/2022
							F1901	Additional Surgery	A070504	Premature Discharge of Battery	23/08/2022	23/08/2022
									A071102	Failure to Interrogate	23/08/2022	23/08/2022
									A071204	No Pacing	23/08/2022	23/08/2022
									A1305	Wireless Communication Problem	23/08/2022	23/08/2022
					CT983	Surgical invasive	E0601	Arrhythmia	A070504	Premature Discharge of Battery	23/08/2022	23/08/2022
									A071102	Failure to Interrogate	23/08/2022	23/08/2022
									A071204	No Pacing	23/08/2022	23/08/2022
									A1305	Wireless Communication Problem	23/08/2022	23/08/2022
							E2311	Discomfort	A070504	Premature Discharge of Battery	23/08/2022	23/08/2022
									A071102	Failure to Interrogate	23/08/2022	23/08/2022
									A071204	No Pacing	23/08/2022	23/08/2022
									A1305	Wireless Communication Problem	23/08/2022	23/08/2022
							E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	23/08/2022	23/08/2022
									A071102	Failure to Interrogate	23/08/2022	23/08/2022
									A071204	No Pacing	23/08/2022	23/08/2022
									A1305	Wireless Communication Problem	23/08/2022	23/08/2022
							F1901	Additional Surgery	A070504	Premature Discharge of Battery	23/08/2022	23/08/2022
									A071102	Failure to Interrogate	23/08/2022	23/08/2022
									A071204	No Pacing	23/08/2022	23/08/2022
									A1305	Wireless Communication Problem	23/08/2022	23/08/2022
7576	2022	All other reportable incidents	47263	Cardiac resynchronization therapy implantable pacemaker	CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A0908	Incorrect, Inadequate or Imprecise Result or Readings	22/08/2022	22/08/2022
							F08	Hospitalization or Prolonged Hospitalization	A0908	Incorrect, Inadequate or Imprecise Result or Readings	22/08/2022	22/08/2022
							F1905	Device Revision or Replacement	A0908	Incorrect, Inadequate or Imprecise Result or Readings	22/08/2022	22/08/2022
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A0908	Incorrect, Inadequate or Imprecise Result or Readings	22/08/2022	22/08/2022
							F08	Hospitalization or Prolonged Hospitalization	A0908	Incorrect, Inadequate or Imprecise Result or Readings	22/08/2022	22/08/2022
							F1905	Device Revision or Replacement	A0908	Incorrect, Inadequate or Imprecise Result or Readings	22/08/2022	22/08/2022
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A0908	Incorrect, Inadequate or Imprecise Result or Readings	22/08/2022	22/08/2022
							F08	Hospitalization or Prolonged Hospitalization	A0908	Incorrect, Inadequate or Imprecise Result or Readings	22/08/2022	22/08/2022
							F1905	Device Revision or Replacement	A0908	Incorrect, Inadequate or Imprecise Result or Readings	22/08/2022	22/08/2022
7577	2022	All other reportable incidents	47267	Single-chamber implantable pacemaker, rate-responsive	CT141	Electrodes/Leads and associated devices	E011903	Syncope/Fainting	A070909	Oversensing	26/08/2022	26/08/2022
									A0712	Pacing Problem	26/08/2022	26/08/2022
									A0908	Incorrect, Inadequate or Imprecise Result or Readings	26/08/2022	26/08/2022
									A090801	Signal Artifact/Noise	26/08/2022	26/08/2022
							E060104	Bradycardia	A070909	Oversensing	26/08/2022	26/08/2022
									A0712	Pacing Problem	26/08/2022	26/08/2022
									A0908	Incorrect, Inadequate or Imprecise Result or Readings	26/08/2022	26/08/2022
									A090801	Signal Artifact/Noise	26/08/2022	26/08/2022
							E2007	Fall	A070909	Oversensing	26/08/2022	26/08/2022
									A0712	Pacing Problem	26/08/2022	26/08/2022

								A0908	Incorrect, Inadequate or Imprecise Result or Readings	26/08/2022	26/08/2022
								A090801	Signal Artifact/Noise	26/08/2022	26/08/2022
							E2403	No Clinical Signs, Symptoms or Conditions	A070909	Oversensing	26/08/2022	26/08/2022
									A0712	Pacing Problem	26/08/2022	26/08/2022
									A0908	Incorrect, Inadequate or Imprecise Result or Readings	26/08/2022	26/08/2022
									A090801	Signal Artifact/Noise	26/08/2022	26/08/2022
							F08	Hospitalization or Prolonged Hospitalization	A070909	Oversensing	26/08/2022	26/08/2022
									A0712	Pacing Problem	26/08/2022	26/08/2022
									A0908	Incorrect, Inadequate or Imprecise Result or Readings	26/08/2022	26/08/2022
									A090801	Signal Artifact/Noise	26/08/2022	26/08/2022
							F1901	Additional Surgery	A070909	Oversensing	26/08/2022	26/08/2022
									A0712	Pacing Problem	26/08/2022	26/08/2022
									A0908	Incorrect, Inadequate or Imprecise Result or Readings	26/08/2022	26/08/2022
									A090801	Signal Artifact/Noise	26/08/2022	26/08/2022
							F1905	Device Revision or Replacement	A070909	Oversensing	26/08/2022	26/08/2022
									A0712	Pacing Problem	26/08/2022	26/08/2022
									A0908	Incorrect, Inadequate or Imprecise Result or Readings	26/08/2022	26/08/2022
									A090801	Signal Artifact/Noise	26/08/2022	26/08/2022
					CT145	Cardiovascular devices	E011903	Syncope/Fainting	A070909	Oversensing	26/08/2022	26/08/2022
									A0712	Pacing Problem	26/08/2022	26/08/2022
									A0908	Incorrect, Inadequate or Imprecise Result or Readings	26/08/2022	26/08/2022
									A090801	Signal Artifact/Noise	26/08/2022	26/08/2022
							E060104	Bradycardia	A070909	Oversensing	26/08/2022	26/08/2022
									A0712	Pacing Problem	26/08/2022	26/08/2022
									A0908	Incorrect, Inadequate or Imprecise Result or Readings	26/08/2022	26/08/2022
									A090801	Signal Artifact/Noise	26/08/2022	26/08/2022
							E2007	Fall	A070909	Oversensing	26/08/2022	26/08/2022
									A0712	Pacing Problem	26/08/2022	26/08/2022
									A0908	Incorrect, Inadequate or Imprecise Result or Readings	26/08/2022	26/08/2022
									A090801	Signal Artifact/Noise	26/08/2022	26/08/2022
							E2403	No Clinical Signs, Symptoms or Conditions	A070909	Oversensing	26/08/2022	26/08/2022
									A0712	Pacing Problem	26/08/2022	26/08/2022
									A0908	Incorrect, Inadequate or Imprecise Result or Readings	26/08/2022	26/08/2022
									A090801	Signal Artifact/Noise	26/08/2022	26/08/2022
							F08	Hospitalization or Prolonged Hospitalization	A070909	Oversensing	26/08/2022	26/08/2022
									A0712	Pacing Problem	26/08/2022	26/08/2022
									A0908	Incorrect, Inadequate or Imprecise Result or Readings	26/08/2022	26/08/2022
									A090801	Signal Artifact/Noise	26/08/2022	26/08/2022
							F1901	Additional Surgery	A070909	Oversensing	26/08/2022	26/08/2022
									A0712	Pacing Problem	26/08/2022	26/08/2022
									A0908	Incorrect, Inadequate or Imprecise Result or Readings	26/08/2022	26/08/2022
									A090801	Signal Artifact/Noise	26/08/2022	26/08/2022
							F1905	Device Revision or Replacement	A070909	Oversensing	26/08/2022	26/08/2022
									A0712	Pacing Problem	26/08/2022	26/08/2022
									A0908	Incorrect, Inadequate or Imprecise Result or Readings	26/08/2022	26/08/2022
									A090801	Signal Artifact/Noise	26/08/2022	26/08/2022
					CT1684	Cardiac pacemakers and associated devices	E011903	Syncope/Fainting	A070909	Oversensing	26/08/2022	26/08/2022
									A0712	Pacing Problem	26/08/2022	26/08/2022
									A0908	Incorrect, Inadequate or Imprecise Result or Readings	26/08/2022	26/08/2022
									A090801	Signal Artifact/Noise	26/08/2022	26/08/2022
							E060104	Bradycardia	A070909	Oversensing	26/08/2022	26/08/2022
									A0712	Pacing Problem	26/08/2022	26/08/2022
									A0908	Incorrect, Inadequate or Imprecise Result or Readings	26/08/2022	26/08/2022
									A090801	Signal Artifact/Noise	26/08/2022	26/08/2022
							E2007	Fall	A070909	Oversensing	26/08/2022	26/08/2022
									A0712	Pacing Problem	26/08/2022	26/08/2022
									A0908	Incorrect, Inadequate or Imprecise Result or Readings	26/08/2022	26/08/2022
									A090801	Signal Artifact/Noise	26/08/2022	26/08/2022
							E2403	No Clinical Signs, Symptoms or Conditions	A070909	Oversensing	26/08/2022	26/08/2022
									A0712	Pacing Problem	26/08/2022	26/08/2022
									A0908	Incorrect, Inadequate or Imprecise Result or Readings	26/08/2022	26/08/2022
									A090801	Signal Artifact/Noise	26/08/2022	26/08/2022
							F08	Hospitalization or Prolonged Hospitalization	A070909	Oversensing	26/08/2022	26/08/2022
									A0712	Pacing Problem	26/08/2022	26/08/2022
									A0908	Incorrect, Inadequate or Imprecise Result or Readings	26/08/2022	26/08/2022
									A090801	Signal Artifact/Noise	26/08/2022	26/08/2022
							F1901	Additional Surgery	A070909	Oversensing	26/08/2022	26/08/2022
									A0712	Pacing Problem	26/08/2022	26/08/2022
									A0908	Incorrect, Inadequate or Imprecise Result or Readings	26/08/2022	26/08/2022
									A090801	Signal Artifact/Noise	26/08/2022	26/08/2022
							F1905	Device Revision or Replacement	A070909	Oversensing	26/08/2022	26/08/2022
									A0712	Pacing Problem	26/08/2022	26/08/2022
									A0908	Incorrect, Inadequate or Imprecise Result or Readings	26/08/2022	26/08/2022

								A090801	Signal Artifact/Noise	26/08/2022	26/08/2022
					CT1685	Cardiac pacemakers	E011903	Syncope/Fainting	A070909	Oversensing	26/08/2022	26/08/2022
									A0712	Pacing Problem	26/08/2022	26/08/2022
									A0908	Incorrect, Inadequate or Imprecise Result or Readings	26/08/2022	26/08/2022
									A090801	Signal Artifact/Noise	26/08/2022	26/08/2022
							E060104	Bradycardia	A070909	Oversensing	26/08/2022	26/08/2022
									A0712	Pacing Problem	26/08/2022	26/08/2022
									A0908	Incorrect, Inadequate or Imprecise Result or Readings	26/08/2022	26/08/2022
									A090801	Signal Artifact/Noise	26/08/2022	26/08/2022
							E2007	Fall	A070909	Oversensing	26/08/2022	26/08/2022
									A0712	Pacing Problem	26/08/2022	26/08/2022
									A0908	Incorrect, Inadequate or Imprecise Result or Readings	26/08/2022	26/08/2022
									A090801	Signal Artifact/Noise	26/08/2022	26/08/2022
							E2403	No Clinical Signs, Symptoms or Conditions	A070909	Oversensing	26/08/2022	26/08/2022
									A0712	Pacing Problem	26/08/2022	26/08/2022
									A0908	Incorrect, Inadequate or Imprecise Result or Readings	26/08/2022	26/08/2022
									A090801	Signal Artifact/Noise	26/08/2022	26/08/2022
							F08	Hospitalization or Prolonged Hospitalization	A070909	Oversensing	26/08/2022	26/08/2022
									A0712	Pacing Problem	26/08/2022	26/08/2022
									A0908	Incorrect, Inadequate or Imprecise Result or Readings	26/08/2022	26/08/2022
									A090801	Signal Artifact/Noise	26/08/2022	26/08/2022
							F1901	Additional Surgery	A070909	Oversensing	26/08/2022	26/08/2022
									A0712	Pacing Problem	26/08/2022	26/08/2022
									A0908	Incorrect, Inadequate or Imprecise Result or Readings	26/08/2022	26/08/2022
									A090801	Signal Artifact/Noise	26/08/2022	26/08/2022
							F1905	Device Revision or Replacement	A070909	Oversensing	26/08/2022	26/08/2022
									A0712	Pacing Problem	26/08/2022	26/08/2022
									A0908	Incorrect, Inadequate or Imprecise Result or Readings	26/08/2022	26/08/2022
									A090801	Signal Artifact/Noise	26/08/2022	26/08/2022
					CT233	Surgical	E011903	Syncope/Fainting	A070909	Oversensing	26/08/2022	26/08/2022
									A0712	Pacing Problem	26/08/2022	26/08/2022
									A0908	Incorrect, Inadequate or Imprecise Result or Readings	26/08/2022	26/08/2022
									A090801	Signal Artifact/Noise	26/08/2022	26/08/2022
							E060104	Bradycardia	A070909	Oversensing	26/08/2022	26/08/2022
									A0712	Pacing Problem	26/08/2022	26/08/2022
									A0908	Incorrect, Inadequate or Imprecise Result or Readings	26/08/2022	26/08/2022
									A090801	Signal Artifact/Noise	26/08/2022	26/08/2022
							E2007	Fall	A070909	Oversensing	26/08/2022	26/08/2022
									A0712	Pacing Problem	26/08/2022	26/08/2022
									A0908	Incorrect, Inadequate or Imprecise Result or Readings	26/08/2022	26/08/2022
									A090801	Signal Artifact/Noise	26/08/2022	26/08/2022
							E2403	No Clinical Signs, Symptoms or Conditions	A070909	Oversensing	26/08/2022	26/08/2022
									A0712	Pacing Problem	26/08/2022	26/08/2022
									A0908	Incorrect, Inadequate or Imprecise Result or Readings	26/08/2022	26/08/2022
									A090801	Signal Artifact/Noise	26/08/2022	26/08/2022
							F08	Hospitalization or Prolonged Hospitalization	A070909	Oversensing	26/08/2022	26/08/2022
									A0712	Pacing Problem	26/08/2022	26/08/2022
									A0908	Incorrect, Inadequate or Imprecise Result or Readings	26/08/2022	26/08/2022
									A090801	Signal Artifact/Noise	26/08/2022	26/08/2022
							F1901	Additional Surgery	A070909	Oversensing	26/08/2022	26/08/2022
									A0712	Pacing Problem	26/08/2022	26/08/2022
									A0908	Incorrect, Inadequate or Imprecise Result or Readings	26/08/2022	26/08/2022
									A090801	Signal Artifact/Noise	26/08/2022	26/08/2022
							F1905	Device Revision or Replacement	A070909	Oversensing	26/08/2022	26/08/2022
									A0712	Pacing Problem	26/08/2022	26/08/2022
									A0908	Incorrect, Inadequate or Imprecise Result or Readings	26/08/2022	26/08/2022
									A090801	Signal Artifact/Noise	26/08/2022	26/08/2022
					CT2406	Implantable	E011903	Syncope/Fainting	A070909	Oversensing	26/08/2022	26/08/2022
									A0712	Pacing Problem	26/08/2022	26/08/2022
									A0908	Incorrect, Inadequate or Imprecise Result or Readings	26/08/2022	26/08/2022
									A090801	Signal Artifact/Noise	26/08/2022	26/08/2022
							E060104	Bradycardia	A070909	Oversensing	26/08/2022	26/08/2022
									A0712	Pacing Problem	26/08/2022	26/08/2022
									A0908	Incorrect, Inadequate or Imprecise Result or Readings	26/08/2022	26/08/2022
									A090801	Signal Artifact/Noise	26/08/2022	26/08/2022
							E2007	Fall	A070909	Oversensing	26/08/2022	26/08/2022
									A0712	Pacing Problem	26/08/2022	26/08/2022
									A0908	Incorrect, Inadequate or Imprecise Result or Readings	26/08/2022	26/08/2022
									A090801	Signal Artifact/Noise	26/08/2022	26/08/2022
							E2403	No Clinical Signs, Symptoms or Conditions	A070909	Oversensing	26/08/2022	26/08/2022
									A0712	Pacing Problem	26/08/2022	26/08/2022
									A0908	Incorrect, Inadequate or Imprecise Result or Readings	26/08/2022	26/08/2022
									A090801	Signal Artifact/Noise	26/08/2022	26/08/2022
							F08	Hospitalization or Prolonged Hospitalization	A070909	Oversensing	26/08/2022	26/08/2022

								A0712	Pacing Problem	26/08/2022	26/08/2022
								A0908	Incorrect, Inadequate or Imprecise Result or Readings	26/08/2022	26/08/2022
								A090801	Signal Artifact/Noise	26/08/2022	26/08/2022
							F1901	Additional Surgery	A070909	Oversensing	26/08/2022	26/08/2022
									A0712	Pacing Problem	26/08/2022	26/08/2022
									A0908	Incorrect, Inadequate or Imprecise Result or Readings	26/08/2022	26/08/2022
									A090801	Signal Artifact/Noise	26/08/2022	26/08/2022
							F1905	Device Revision or Replacement	A070909	Oversensing	26/08/2022	26/08/2022
									A0712	Pacing Problem	26/08/2022	26/08/2022
									A0908	Incorrect, Inadequate or Imprecise Result or Readings	26/08/2022	26/08/2022
									A090801	Signal Artifact/Noise	26/08/2022	26/08/2022
					CT2407	Active implantable	E011903	Syncope/Fainting	A070909	Oversensing	26/08/2022	26/08/2022
									A0712	Pacing Problem	26/08/2022	26/08/2022
									A0908	Incorrect, Inadequate or Imprecise Result or Readings	26/08/2022	26/08/2022
									A090801	Signal Artifact/Noise	26/08/2022	26/08/2022
							E060104	Bradycardia	A070909	Oversensing	26/08/2022	26/08/2022
									A0712	Pacing Problem	26/08/2022	26/08/2022
									A0908	Incorrect, Inadequate or Imprecise Result or Readings	26/08/2022	26/08/2022
									A090801	Signal Artifact/Noise	26/08/2022	26/08/2022
							E2007	Fall	A070909	Oversensing	26/08/2022	26/08/2022
									A0712	Pacing Problem	26/08/2022	26/08/2022
									A0908	Incorrect, Inadequate or Imprecise Result or Readings	26/08/2022	26/08/2022
									A090801	Signal Artifact/Noise	26/08/2022	26/08/2022
							E2403	No Clinical Signs, Symptoms or Conditions	A070909	Oversensing	26/08/2022	26/08/2022
									A0712	Pacing Problem	26/08/2022	26/08/2022
									A0908	Incorrect, Inadequate or Imprecise Result or Readings	26/08/2022	26/08/2022
									A090801	Signal Artifact/Noise	26/08/2022	26/08/2022
							F08	Hospitalization or Prolonged Hospitalization	A070909	Oversensing	26/08/2022	26/08/2022
									A0712	Pacing Problem	26/08/2022	26/08/2022
									A0908	Incorrect, Inadequate or Imprecise Result or Readings	26/08/2022	26/08/2022
									A090801	Signal Artifact/Noise	26/08/2022	26/08/2022
							F1901	Additional Surgery	A070909	Oversensing	26/08/2022	26/08/2022
									A0712	Pacing Problem	26/08/2022	26/08/2022
									A0908	Incorrect, Inadequate or Imprecise Result or Readings	26/08/2022	26/08/2022
									A090801	Signal Artifact/Noise	26/08/2022	26/08/2022
							F1905	Device Revision or Replacement	A070909	Oversensing	26/08/2022	26/08/2022
									A0712	Pacing Problem	26/08/2022	26/08/2022
									A0908	Incorrect, Inadequate or Imprecise Result or Readings	26/08/2022	26/08/2022
									A090801	Signal Artifact/Noise	26/08/2022	26/08/2022
					CT257	Implantable cardiac pacemakers	E011903	Syncope/Fainting	A070909	Oversensing	26/08/2022	26/08/2022
									A0712	Pacing Problem	26/08/2022	26/08/2022
									A0908	Incorrect, Inadequate or Imprecise Result or Readings	26/08/2022	26/08/2022
									A090801	Signal Artifact/Noise	26/08/2022	26/08/2022
							E060104	Bradycardia	A070909	Oversensing	26/08/2022	26/08/2022
									A0712	Pacing Problem	26/08/2022	26/08/2022
									A0908	Incorrect, Inadequate or Imprecise Result or Readings	26/08/2022	26/08/2022
									A090801	Signal Artifact/Noise	26/08/2022	26/08/2022
							E2007	Fall	A070909	Oversensing	26/08/2022	26/08/2022
									A0712	Pacing Problem	26/08/2022	26/08/2022
									A0908	Incorrect, Inadequate or Imprecise Result or Readings	26/08/2022	26/08/2022
									A090801	Signal Artifact/Noise	26/08/2022	26/08/2022
							E2403	No Clinical Signs, Symptoms or Conditions	A070909	Oversensing	26/08/2022	26/08/2022
									A0712	Pacing Problem	26/08/2022	26/08/2022
									A0908	Incorrect, Inadequate or Imprecise Result or Readings	26/08/2022	26/08/2022
									A090801	Signal Artifact/Noise	26/08/2022	26/08/2022
							F08	Hospitalization or Prolonged Hospitalization	A070909	Oversensing	26/08/2022	26/08/2022
									A0712	Pacing Problem	26/08/2022	26/08/2022
									A0908	Incorrect, Inadequate or Imprecise Result or Readings	26/08/2022	26/08/2022
									A090801	Signal Artifact/Noise	26/08/2022	26/08/2022
							F1901	Additional Surgery	A070909	Oversensing	26/08/2022	26/08/2022
									A0712	Pacing Problem	26/08/2022	26/08/2022
									A0908	Incorrect, Inadequate or Imprecise Result or Readings	26/08/2022	26/08/2022
									A090801	Signal Artifact/Noise	26/08/2022	26/08/2022
							F1905	Device Revision or Replacement	A070909	Oversensing	26/08/2022	26/08/2022
									A0712	Pacing Problem	26/08/2022	26/08/2022
									A0908	Incorrect, Inadequate or Imprecise Result or Readings	26/08/2022	26/08/2022
									A090801	Signal Artifact/Noise	26/08/2022	26/08/2022
					CT2727	Active implant devices/ancillaries	E011903	Syncope/Fainting	A070909	Oversensing	26/08/2022	26/08/2022
									A0712	Pacing Problem	26/08/2022	26/08/2022
									A0908	Incorrect, Inadequate or Imprecise Result or Readings	26/08/2022	26/08/2022
									A090801	Signal Artifact/Noise	26/08/2022	26/08/2022
							E060104	Bradycardia	A070909	Oversensing	26/08/2022	26/08/2022
									A0712	Pacing Problem	26/08/2022	26/08/2022

								A0908	Incorrect, Inadequate or Imprecise Result or Readings	26/08/2022	26/08/2022
								A090801	Signal Artifact/Noise	26/08/2022	26/08/2022
							E2007	Fall	A070909	Oversensing	26/08/2022	26/08/2022
									A0712	Pacing Problem	26/08/2022	26/08/2022
									A0908	Incorrect, Inadequate or Imprecise Result or Readings	26/08/2022	26/08/2022
									A090801	Signal Artifact/Noise	26/08/2022	26/08/2022
							E2403	No Clinical Signs, Symptoms or Conditions	A070909	Oversensing	26/08/2022	26/08/2022
									A0712	Pacing Problem	26/08/2022	26/08/2022
									A0908	Incorrect, Inadequate or Imprecise Result or Readings	26/08/2022	26/08/2022
									A090801	Signal Artifact/Noise	26/08/2022	26/08/2022
							F08	Hospitalization or Prolonged Hospitalization	A070909	Oversensing	26/08/2022	26/08/2022
									A0712	Pacing Problem	26/08/2022	26/08/2022
									A0908	Incorrect, Inadequate or Imprecise Result or Readings	26/08/2022	26/08/2022
									A090801	Signal Artifact/Noise	26/08/2022	26/08/2022
							F1901	Additional Surgery	A070909	Oversensing	26/08/2022	26/08/2022
									A0712	Pacing Problem	26/08/2022	26/08/2022
									A0908	Incorrect, Inadequate or Imprecise Result or Readings	26/08/2022	26/08/2022
									A090801	Signal Artifact/Noise	26/08/2022	26/08/2022
							F1905	Device Revision or Replacement	A070909	Oversensing	26/08/2022	26/08/2022
									A0712	Pacing Problem	26/08/2022	26/08/2022
									A0908	Incorrect, Inadequate or Imprecise Result or Readings	26/08/2022	26/08/2022
									A090801	Signal Artifact/Noise	26/08/2022	26/08/2022
						CT2750	Implantable cardiac electrical stimulators and associated devices	E011903	Syncope/Fainting	A070909	Oversensing	26/08/2022	26/08/2022
										A0712	Pacing Problem	26/08/2022	26/08/2022
										A0908	Incorrect, Inadequate or Imprecise Result or Readings	26/08/2022	26/08/2022
										A090801	Signal Artifact/Noise	26/08/2022	26/08/2022
								E060104	Bradycardia	A070909	Oversensing	26/08/2022	26/08/2022
										A0712	Pacing Problem	26/08/2022	26/08/2022
										A0908	Incorrect, Inadequate or Imprecise Result or Readings	26/08/2022	26/08/2022
										A090801	Signal Artifact/Noise	26/08/2022	26/08/2022
								E2007	Fall	A070909	Oversensing	26/08/2022	26/08/2022
										A0712	Pacing Problem	26/08/2022	26/08/2022
										A0908	Incorrect, Inadequate or Imprecise Result or Readings	26/08/2022	26/08/2022
										A090801	Signal Artifact/Noise	26/08/2022	26/08/2022
								E2403	No Clinical Signs, Symptoms or Conditions	A070909	Oversensing	26/08/2022	26/08/2022
										A0712	Pacing Problem	26/08/2022	26/08/2022
										A0908	Incorrect, Inadequate or Imprecise Result or Readings	26/08/2022	26/08/2022
										A090801	Signal Artifact/Noise	26/08/2022	26/08/2022
								F08	Hospitalization or Prolonged Hospitalization	A070909	Oversensing	26/08/2022	26/08/2022
										A0712	Pacing Problem	26/08/2022	26/08/2022
										A0908	Incorrect, Inadequate or Imprecise Result or Readings	26/08/2022	26/08/2022
										A090801	Signal Artifact/Noise	26/08/2022	26/08/2022
								F1901	Additional Surgery	A070909	Oversensing	26/08/2022	26/08/2022
										A0712	Pacing Problem	26/08/2022	26/08/2022
										A0908	Incorrect, Inadequate or Imprecise Result or Readings	26/08/2022	26/08/2022
										A090801	Signal Artifact/Noise	26/08/2022	26/08/2022
								F1905	Device Revision or Replacement	A070909	Oversensing	26/08/2022	26/08/2022
										A0712	Pacing Problem	26/08/2022	26/08/2022
										A0908	Incorrect, Inadequate or Imprecise Result or Readings	26/08/2022	26/08/2022
										A090801	Signal Artifact/Noise	26/08/2022	26/08/2022
						CT2866	Implantable cardiac electrical stimulators	E011903	Syncope/Fainting	A070909	Oversensing	26/08/2022	26/08/2022
										A0712	Pacing Problem	26/08/2022	26/08/2022
										A0908	Incorrect, Inadequate or Imprecise Result or Readings	26/08/2022	26/08/2022
										A090801	Signal Artifact/Noise	26/08/2022	26/08/2022
								E060104	Bradycardia	A070909	Oversensing	26/08/2022	26/08/2022
										A0712	Pacing Problem	26/08/2022	26/08/2022
										A0908	Incorrect, Inadequate or Imprecise Result or Readings	26/08/2022	26/08/2022
										A090801	Signal Artifact/Noise	26/08/2022	26/08/2022
								E2007	Fall	A070909	Oversensing	26/08/2022	26/08/2022
										A0712	Pacing Problem	26/08/2022	26/08/2022
										A0908	Incorrect, Inadequate or Imprecise Result or Readings	26/08/2022	26/08/2022
										A090801	Signal Artifact/Noise	26/08/2022	26/08/2022
								E2403	No Clinical Signs, Symptoms or Conditions	A070909	Oversensing	26/08/2022	26/08/2022
										A0712	Pacing Problem	26/08/2022	26/08/2022
										A0908	Incorrect, Inadequate or Imprecise Result or Readings	26/08/2022	26/08/2022
										A090801	Signal Artifact/Noise	26/08/2022	26/08/2022
								F08	Hospitalization or Prolonged Hospitalization	A070909	Oversensing	26/08/2022	26/08/2022
										A0712	Pacing Problem	26/08/2022	26/08/2022
										A0908	Incorrect, Inadequate or Imprecise Result or Readings	26/08/2022	26/08/2022
										A090801	Signal Artifact/Noise	26/08/2022	26/08/2022
								F1901	Additional Surgery	A070909	Oversensing	26/08/2022	26/08/2022
										A0712	Pacing Problem	26/08/2022	26/08/2022
										A0908	Incorrect, Inadequate or Imprecise Result or Readings	26/08/2022	26/08/2022

								A090801	Signal Artifact/Noise	26/08/2022	26/08/2022
						F1905	Device Revision or Replacement	A070909	Oversensing	26/08/2022	26/08/2022
								A0712	Pacing Problem	26/08/2022	26/08/2022
								A0908	Incorrect, Inadequate or Imprecise Result or Readings	26/08/2022	26/08/2022
								A090801	Signal Artifact/Noise	26/08/2022	26/08/2022
					CT2985	Active	E011903	Syncope/Fainting	A070909	Oversensing	26/08/2022	26/08/2022
									A0712	Pacing Problem	26/08/2022	26/08/2022
									A0908	Incorrect, Inadequate or Imprecise Result or Readings	26/08/2022	26/08/2022
									A090801	Signal Artifact/Noise	26/08/2022	26/08/2022
							E060104	Bradycardia	A070909	Oversensing	26/08/2022	26/08/2022
									A0712	Pacing Problem	26/08/2022	26/08/2022
									A0908	Incorrect, Inadequate or Imprecise Result or Readings	26/08/2022	26/08/2022
									A090801	Signal Artifact/Noise	26/08/2022	26/08/2022
							E2007	Fall	A070909	Oversensing	26/08/2022	26/08/2022
									A0712	Pacing Problem	26/08/2022	26/08/2022
									A0908	Incorrect, Inadequate or Imprecise Result or Readings	26/08/2022	26/08/2022
									A090801	Signal Artifact/Noise	26/08/2022	26/08/2022
							E2403	No Clinical Signs, Symptoms or Conditions	A070909	Oversensing	26/08/2022	26/08/2022
									A0712	Pacing Problem	26/08/2022	26/08/2022
									A0908	Incorrect, Inadequate or Imprecise Result or Readings	26/08/2022	26/08/2022
									A090801	Signal Artifact/Noise	26/08/2022	26/08/2022
							F08	Hospitalization or Prolonged Hospitalization	A070909	Oversensing	26/08/2022	26/08/2022
									A0712	Pacing Problem	26/08/2022	26/08/2022
									A0908	Incorrect, Inadequate or Imprecise Result or Readings	26/08/2022	26/08/2022
									A090801	Signal Artifact/Noise	26/08/2022	26/08/2022
							F1901	Additional Surgery	A070909	Oversensing	26/08/2022	26/08/2022
									A0712	Pacing Problem	26/08/2022	26/08/2022
									A0908	Incorrect, Inadequate or Imprecise Result or Readings	26/08/2022	26/08/2022
									A090801	Signal Artifact/Noise	26/08/2022	26/08/2022
							F1905	Device Revision or Replacement	A070909	Oversensing	26/08/2022	26/08/2022
									A0712	Pacing Problem	26/08/2022	26/08/2022
									A0908	Incorrect, Inadequate or Imprecise Result or Readings	26/08/2022	26/08/2022
									A090801	Signal Artifact/Noise	26/08/2022	26/08/2022
					CT3001	Implantable metal	E011903	Syncope/Fainting	A070909	Oversensing	26/08/2022	26/08/2022
									A0712	Pacing Problem	26/08/2022	26/08/2022
									A0908	Incorrect, Inadequate or Imprecise Result or Readings	26/08/2022	26/08/2022
									A090801	Signal Artifact/Noise	26/08/2022	26/08/2022
							E060104	Bradycardia	A070909	Oversensing	26/08/2022	26/08/2022
									A0712	Pacing Problem	26/08/2022	26/08/2022
									A0908	Incorrect, Inadequate or Imprecise Result or Readings	26/08/2022	26/08/2022
									A090801	Signal Artifact/Noise	26/08/2022	26/08/2022
							E2007	Fall	A070909	Oversensing	26/08/2022	26/08/2022
									A0712	Pacing Problem	26/08/2022	26/08/2022
									A0908	Incorrect, Inadequate or Imprecise Result or Readings	26/08/2022	26/08/2022
									A090801	Signal Artifact/Noise	26/08/2022	26/08/2022
							E2403	No Clinical Signs, Symptoms or Conditions	A070909	Oversensing	26/08/2022	26/08/2022
									A0712	Pacing Problem	26/08/2022	26/08/2022
									A0908	Incorrect, Inadequate or Imprecise Result or Readings	26/08/2022	26/08/2022
									A090801	Signal Artifact/Noise	26/08/2022	26/08/2022
							F08	Hospitalization or Prolonged Hospitalization	A070909	Oversensing	26/08/2022	26/08/2022
									A0712	Pacing Problem	26/08/2022	26/08/2022
									A0908	Incorrect, Inadequate or Imprecise Result or Readings	26/08/2022	26/08/2022
									A090801	Signal Artifact/Noise	26/08/2022	26/08/2022
							F1901	Additional Surgery	A070909	Oversensing	26/08/2022	26/08/2022
									A0712	Pacing Problem	26/08/2022	26/08/2022
									A0908	Incorrect, Inadequate or Imprecise Result or Readings	26/08/2022	26/08/2022
									A090801	Signal Artifact/Noise	26/08/2022	26/08/2022
							F1905	Device Revision or Replacement	A070909	Oversensing	26/08/2022	26/08/2022
									A0712	Pacing Problem	26/08/2022	26/08/2022
									A0908	Incorrect, Inadequate or Imprecise Result or Readings	26/08/2022	26/08/2022
									A090801	Signal Artifact/Noise	26/08/2022	26/08/2022
					CT3002	Implantable synthetic polymer	E011903	Syncope/Fainting	A070909	Oversensing	26/08/2022	26/08/2022
									A0712	Pacing Problem	26/08/2022	26/08/2022
									A0908	Incorrect, Inadequate or Imprecise Result or Readings	26/08/2022	26/08/2022
									A090801	Signal Artifact/Noise	26/08/2022	26/08/2022
							E060104	Bradycardia	A070909	Oversensing	26/08/2022	26/08/2022
									A0712	Pacing Problem	26/08/2022	26/08/2022
									A0908	Incorrect, Inadequate or Imprecise Result or Readings	26/08/2022	26/08/2022
									A090801	Signal Artifact/Noise	26/08/2022	26/08/2022
							E2007	Fall	A070909	Oversensing	26/08/2022	26/08/2022
									A0712	Pacing Problem	26/08/2022	26/08/2022
									A0908	Incorrect, Inadequate or Imprecise Result or Readings	26/08/2022	26/08/2022
									A090801	Signal Artifact/Noise	26/08/2022	26/08/2022
							E2403	No Clinical Signs, Symptoms or Conditions	A070909	Oversensing	26/08/2022	26/08/2022

								A0712	Pacing Problem	26/08/2022	26/08/2022
								A0908	Incorrect, Inadequate or Imprecise Result or Readings	26/08/2022	26/08/2022
								A090801	Signal Artifact/Noise	26/08/2022	26/08/2022
							F08	Hospitalization or Prolonged Hospitalization	A070909	Oversensing	26/08/2022	26/08/2022
									A0712	Pacing Problem	26/08/2022	26/08/2022
									A0908	Incorrect, Inadequate or Imprecise Result or Readings	26/08/2022	26/08/2022
									A090801	Signal Artifact/Noise	26/08/2022	26/08/2022
							F1901	Additional Surgery	A070909	Oversensing	26/08/2022	26/08/2022
									A0712	Pacing Problem	26/08/2022	26/08/2022
									A0908	Incorrect, Inadequate or Imprecise Result or Readings	26/08/2022	26/08/2022
									A090801	Signal Artifact/Noise	26/08/2022	26/08/2022
							F1905	Device Revision or Replacement	A070909	Oversensing	26/08/2022	26/08/2022
									A0712	Pacing Problem	26/08/2022	26/08/2022
									A0908	Incorrect, Inadequate or Imprecise Result or Readings	26/08/2022	26/08/2022
									A090801	Signal Artifact/Noise	26/08/2022	26/08/2022
					CT3006	Implantable inorganic material	E011903	Syncope/Fainting	A070909	Oversensing	26/08/2022	26/08/2022
									A0712	Pacing Problem	26/08/2022	26/08/2022
									A0908	Incorrect, Inadequate or Imprecise Result or Readings	26/08/2022	26/08/2022
									A090801	Signal Artifact/Noise	26/08/2022	26/08/2022
							E060104	Bradycardia	A070909	Oversensing	26/08/2022	26/08/2022
									A0712	Pacing Problem	26/08/2022	26/08/2022
									A0908	Incorrect, Inadequate or Imprecise Result or Readings	26/08/2022	26/08/2022
									A090801	Signal Artifact/Noise	26/08/2022	26/08/2022
							E2007	Fall	A070909	Oversensing	26/08/2022	26/08/2022
									A0712	Pacing Problem	26/08/2022	26/08/2022
									A0908	Incorrect, Inadequate or Imprecise Result or Readings	26/08/2022	26/08/2022
									A090801	Signal Artifact/Noise	26/08/2022	26/08/2022
							E2403	No Clinical Signs, Symptoms or Conditions	A070909	Oversensing	26/08/2022	26/08/2022
									A0712	Pacing Problem	26/08/2022	26/08/2022
									A0908	Incorrect, Inadequate or Imprecise Result or Readings	26/08/2022	26/08/2022
									A090801	Signal Artifact/Noise	26/08/2022	26/08/2022
							F08	Hospitalization or Prolonged Hospitalization	A070909	Oversensing	26/08/2022	26/08/2022
									A0712	Pacing Problem	26/08/2022	26/08/2022
									A0908	Incorrect, Inadequate or Imprecise Result or Readings	26/08/2022	26/08/2022
									A090801	Signal Artifact/Noise	26/08/2022	26/08/2022
							F1901	Additional Surgery	A070909	Oversensing	26/08/2022	26/08/2022
									A0712	Pacing Problem	26/08/2022	26/08/2022
									A0908	Incorrect, Inadequate or Imprecise Result or Readings	26/08/2022	26/08/2022
									A090801	Signal Artifact/Noise	26/08/2022	26/08/2022
							F1905	Device Revision or Replacement	A070909	Oversensing	26/08/2022	26/08/2022
									A0712	Pacing Problem	26/08/2022	26/08/2022
									A0908	Incorrect, Inadequate or Imprecise Result or Readings	26/08/2022	26/08/2022
									A090801	Signal Artifact/Noise	26/08/2022	26/08/2022
					CT314	Electrically powered	E011903	Syncope/Fainting	A070909	Oversensing	26/08/2022	26/08/2022
									A0712	Pacing Problem	26/08/2022	26/08/2022
									A0908	Incorrect, Inadequate or Imprecise Result or Readings	26/08/2022	26/08/2022
									A090801	Signal Artifact/Noise	26/08/2022	26/08/2022
							E060104	Bradycardia	A070909	Oversensing	26/08/2022	26/08/2022
									A0712	Pacing Problem	26/08/2022	26/08/2022
									A0908	Incorrect, Inadequate or Imprecise Result or Readings	26/08/2022	26/08/2022
									A090801	Signal Artifact/Noise	26/08/2022	26/08/2022
							E2007	Fall	A070909	Oversensing	26/08/2022	26/08/2022
									A0712	Pacing Problem	26/08/2022	26/08/2022
									A0908	Incorrect, Inadequate or Imprecise Result or Readings	26/08/2022	26/08/2022
									A090801	Signal Artifact/Noise	26/08/2022	26/08/2022
							E2403	No Clinical Signs, Symptoms or Conditions	A070909	Oversensing	26/08/2022	26/08/2022
									A0712	Pacing Problem	26/08/2022	26/08/2022
									A0908	Incorrect, Inadequate or Imprecise Result or Readings	26/08/2022	26/08/2022
									A090801	Signal Artifact/Noise	26/08/2022	26/08/2022
							F08	Hospitalization or Prolonged Hospitalization	A070909	Oversensing	26/08/2022	26/08/2022
									A0712	Pacing Problem	26/08/2022	26/08/2022
									A0908	Incorrect, Inadequate or Imprecise Result or Readings	26/08/2022	26/08/2022
									A090801	Signal Artifact/Noise	26/08/2022	26/08/2022
							F1901	Additional Surgery	A070909	Oversensing	26/08/2022	26/08/2022
									A0712	Pacing Problem	26/08/2022	26/08/2022
									A0908	Incorrect, Inadequate or Imprecise Result or Readings	26/08/2022	26/08/2022
									A090801	Signal Artifact/Noise	26/08/2022	26/08/2022
							F1905	Device Revision or Replacement	A070909	Oversensing	26/08/2022	26/08/2022
									A0712	Pacing Problem	26/08/2022	26/08/2022
									A0908	Incorrect, Inadequate or Imprecise Result or Readings	26/08/2022	26/08/2022
									A090801	Signal Artifact/Noise	26/08/2022	26/08/2022
					CT321	Long-term surgical invasive	E011903	Syncope/Fainting	A070909	Oversensing	26/08/2022	26/08/2022
									A0712	Pacing Problem	26/08/2022	26/08/2022

								A0908	Incorrect, Inadequate or Imprecise Result or Readings	26/08/2022	26/08/2022
								A090801	Signal Artifact/Noise	26/08/2022	26/08/2022
							E060104	Bradycardia	A070909	Oversensing	26/08/2022	26/08/2022
									A0712	Pacing Problem	26/08/2022	26/08/2022
									A0908	Incorrect, Inadequate or Imprecise Result or Readings	26/08/2022	26/08/2022
									A090801	Signal Artifact/Noise	26/08/2022	26/08/2022
							E2007	Fall	A070909	Oversensing	26/08/2022	26/08/2022
									A0712	Pacing Problem	26/08/2022	26/08/2022
									A0908	Incorrect, Inadequate or Imprecise Result or Readings	26/08/2022	26/08/2022
									A090801	Signal Artifact/Noise	26/08/2022	26/08/2022
							E2403	No Clinical Signs, Symptoms or Conditions	A070909	Oversensing	26/08/2022	26/08/2022
									A0712	Pacing Problem	26/08/2022	26/08/2022
									A0908	Incorrect, Inadequate or Imprecise Result or Readings	26/08/2022	26/08/2022
									A090801	Signal Artifact/Noise	26/08/2022	26/08/2022
							F08	Hospitalization or Prolonged Hospitalization	A070909	Oversensing	26/08/2022	26/08/2022
									A0712	Pacing Problem	26/08/2022	26/08/2022
									A0908	Incorrect, Inadequate or Imprecise Result or Readings	26/08/2022	26/08/2022
									A090801	Signal Artifact/Noise	26/08/2022	26/08/2022
							F1901	Additional Surgery	A070909	Oversensing	26/08/2022	26/08/2022
									A0712	Pacing Problem	26/08/2022	26/08/2022
									A0908	Incorrect, Inadequate or Imprecise Result or Readings	26/08/2022	26/08/2022
									A090801	Signal Artifact/Noise	26/08/2022	26/08/2022
							F1905	Device Revision or Replacement	A070909	Oversensing	26/08/2022	26/08/2022
									A0712	Pacing Problem	26/08/2022	26/08/2022
									A0908	Incorrect, Inadequate or Imprecise Result or Readings	26/08/2022	26/08/2022
									A090801	Signal Artifact/Noise	26/08/2022	26/08/2022
						CT449	Generators and associated devices	E011903	Syncope/Fainting	A070909	Oversensing	26/08/2022	26/08/2022
										A0712	Pacing Problem	26/08/2022	26/08/2022
										A0908	Incorrect, Inadequate or Imprecise Result or Readings	26/08/2022	26/08/2022
										A090801	Signal Artifact/Noise	26/08/2022	26/08/2022
								E060104	Bradycardia	A070909	Oversensing	26/08/2022	26/08/2022
										A0712	Pacing Problem	26/08/2022	26/08/2022
										A0908	Incorrect, Inadequate or Imprecise Result or Readings	26/08/2022	26/08/2022
										A090801	Signal Artifact/Noise	26/08/2022	26/08/2022
								E2007	Fall	A070909	Oversensing	26/08/2022	26/08/2022
										A0712	Pacing Problem	26/08/2022	26/08/2022
										A0908	Incorrect, Inadequate or Imprecise Result or Readings	26/08/2022	26/08/2022
										A090801	Signal Artifact/Noise	26/08/2022	26/08/2022
								E2403	No Clinical Signs, Symptoms or Conditions	A070909	Oversensing	26/08/2022	26/08/2022
										A0712	Pacing Problem	26/08/2022	26/08/2022
										A0908	Incorrect, Inadequate or Imprecise Result or Readings	26/08/2022	26/08/2022
										A090801	Signal Artifact/Noise	26/08/2022	26/08/2022
								F08	Hospitalization or Prolonged Hospitalization	A070909	Oversensing	26/08/2022	26/08/2022
										A0712	Pacing Problem	26/08/2022	26/08/2022
										A0908	Incorrect, Inadequate or Imprecise Result or Readings	26/08/2022	26/08/2022
										A090801	Signal Artifact/Noise	26/08/2022	26/08/2022
								F1901	Additional Surgery	A070909	Oversensing	26/08/2022	26/08/2022
										A0712	Pacing Problem	26/08/2022	26/08/2022
										A0908	Incorrect, Inadequate or Imprecise Result or Readings	26/08/2022	26/08/2022
										A090801	Signal Artifact/Noise	26/08/2022	26/08/2022
								F1905	Device Revision or Replacement	A070909	Oversensing	26/08/2022	26/08/2022
										A0712	Pacing Problem	26/08/2022	26/08/2022
										A0908	Incorrect, Inadequate or Imprecise Result or Readings	26/08/2022	26/08/2022
										A090801	Signal Artifact/Noise	26/08/2022	26/08/2022
						CT981	Single use	E011903	Syncope/Fainting	A070909	Oversensing	26/08/2022	26/08/2022
										A0712	Pacing Problem	26/08/2022	26/08/2022
										A0908	Incorrect, Inadequate or Imprecise Result or Readings	26/08/2022	26/08/2022
										A090801	Signal Artifact/Noise	26/08/2022	26/08/2022
								E060104	Bradycardia	A070909	Oversensing	26/08/2022	26/08/2022
										A0712	Pacing Problem	26/08/2022	26/08/2022
										A0908	Incorrect, Inadequate or Imprecise Result or Readings	26/08/2022	26/08/2022
										A090801	Signal Artifact/Noise	26/08/2022	26/08/2022
								E2007	Fall	A070909	Oversensing	26/08/2022	26/08/2022
										A0712	Pacing Problem	26/08/2022	26/08/2022
										A0908	Incorrect, Inadequate or Imprecise Result or Readings	26/08/2022	26/08/2022
										A090801	Signal Artifact/Noise	26/08/2022	26/08/2022
								E2403	No Clinical Signs, Symptoms or Conditions	A070909	Oversensing	26/08/2022	26/08/2022
										A0712	Pacing Problem	26/08/2022	26/08/2022
										A0908	Incorrect, Inadequate or Imprecise Result or Readings	26/08/2022	26/08/2022
										A090801	Signal Artifact/Noise	26/08/2022	26/08/2022
								F08	Hospitalization or Prolonged Hospitalization	A070909	Oversensing	26/08/2022	26/08/2022
										A0712	Pacing Problem	26/08/2022	26/08/2022
										A0908	Incorrect, Inadequate or Imprecise Result or Readings	26/08/2022	26/08/2022

										A090801	Signal Artifact/Noise	26/08/2022	26/08/2022
							F1901	Additional Surgery	A070909	Oversensing	26/08/2022	26/08/2022
									A0712	Pacing Problem	26/08/2022	26/08/2022
									A0908	Incorrect, Inadequate or Imprecise Result or Readings	26/08/2022	26/08/2022
									A090801	Signal Artifact/Noise	26/08/2022	26/08/2022
							F1905	Device Revision or Replacement	A070909	Oversensing	26/08/2022	26/08/2022
									A0712	Pacing Problem	26/08/2022	26/08/2022
									A0908	Incorrect, Inadequate or Imprecise Result or Readings	26/08/2022	26/08/2022
									A090801	Signal Artifact/Noise	26/08/2022	26/08/2022
					CT983	Surgical invasive	E011903	Syncope/Fainting	A070909	Oversensing	26/08/2022	26/08/2022
									A0712	Pacing Problem	26/08/2022	26/08/2022
									A0908	Incorrect, Inadequate or Imprecise Result or Readings	26/08/2022	26/08/2022
									A090801	Signal Artifact/Noise	26/08/2022	26/08/2022
							E060104	Bradycardia	A070909	Oversensing	26/08/2022	26/08/2022
									A0712	Pacing Problem	26/08/2022	26/08/2022
									A0908	Incorrect, Inadequate or Imprecise Result or Readings	26/08/2022	26/08/2022
									A090801	Signal Artifact/Noise	26/08/2022	26/08/2022
							E2007	Fall	A070909	Oversensing	26/08/2022	26/08/2022
									A0712	Pacing Problem	26/08/2022	26/08/2022
									A0908	Incorrect, Inadequate or Imprecise Result or Readings	26/08/2022	26/08/2022
									A090801	Signal Artifact/Noise	26/08/2022	26/08/2022
							E2403	No Clinical Signs, Symptoms or Conditions	A070909	Oversensing	26/08/2022	26/08/2022
									A0712	Pacing Problem	26/08/2022	26/08/2022
									A0908	Incorrect, Inadequate or Imprecise Result or Readings	26/08/2022	26/08/2022
									A090801	Signal Artifact/Noise	26/08/2022	26/08/2022
							F08	Hospitalization or Prolonged Hospitalization	A070909	Oversensing	26/08/2022	26/08/2022
									A0712	Pacing Problem	26/08/2022	26/08/2022
									A0908	Incorrect, Inadequate or Imprecise Result or Readings	26/08/2022	26/08/2022
									A090801	Signal Artifact/Noise	26/08/2022	26/08/2022
							F1901	Additional Surgery	A070909	Oversensing	26/08/2022	26/08/2022
									A0712	Pacing Problem	26/08/2022	26/08/2022
									A0908	Incorrect, Inadequate or Imprecise Result or Readings	26/08/2022	26/08/2022
									A090801	Signal Artifact/Noise	26/08/2022	26/08/2022
							F1905	Device Revision or Replacement	A070909	Oversensing	26/08/2022	26/08/2022
									A0712	Pacing Problem	26/08/2022	26/08/2022
									A0908	Incorrect, Inadequate or Imprecise Result or Readings	26/08/2022	26/08/2022
									A090801	Signal Artifact/Noise	26/08/2022	26/08/2022
7578	2022	All other reportable incidents	47265	Dual-chamber implantable pacemaker, rate-responsive	CT141	Electrodes/Leads and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT145	Cardiovascular devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT233	Surgical	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT2406	Implantable	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT2407	Active implantable	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT2727	Active implant devices/ancillaries	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT2750	Implantable cardiac electrical stimulators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT2985	Active	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT3001	Implantable metal	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT3002	Implantable synthetic polymer	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT3006	Implantable inorganic material	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT314	Electrically powered	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT321	Long-term surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT449	Generators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT981	Single use	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT983	Surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
7579	2022	All other reportable incidents	47265	Dual-chamber implantable pacemaker, rate-responsive	CT141	Electrodes/Leads and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT145	Cardiovascular devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		

					CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT233	Surgical	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT2406	Implantable	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT2407	Active implantable	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT2727	Active implant devices/ancillaries	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT2750	Implantable cardiac electrical stimulators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT2985	Active	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT3001	Implantable metal	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT3002	Implantable synthetic polymer	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT3006	Implantable inorganic material	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT314	Electrically powered	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT321	Long-term surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT449	Generators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT981	Single use	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT983	Surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
7580	2022	All other reportable incidents	47263	Cardiac resynchronization therapy implantable pacemaker	CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
7581	2022	All other reportable incidents	47270	Cardiac resynchronization therapy implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	05/01/2022	05/01/2022
							F1901	Additional Surgery	A070504	Premature Discharge of Battery	05/01/2022	05/01/2022
7582	2022	All other reportable incidents	47205	Cardiac pulse generator programmer	CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A071102	Failure to Interrogate		
									A0712	Pacing Problem		
									A11	Computer Software Problem		
									A1104	Computer Operating System Problem		
							F26	No Health Consequences or Impact	A071102	Failure to Interrogate		
									A0712	Pacing Problem		
									A11	Computer Software Problem		
									A1104	Computer Operating System Problem		
7583	2022	All other reportable incidents	35223	Endocardial/interventricular septal pacing lead	CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A07	Electrical /Electronic Property Problem		
									A070909	Oversensing		
									A0722	Impedance Problem		
							F26	No Health Consequences or Impact	A07	Electrical /Electronic Property Problem		
									A070909	Oversensing		
									A0722	Impedance Problem		
7584	2022	All other reportable incidents	35852	Single-chamber implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A072201	High Impedance		
							F26	No Health Consequences or Impact	A072201	High Impedance		
7585	2022	All other reportable incidents	36241	Pacemaker/defibrillator lead	CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A072201	High Impedance		
							F26	No Health Consequences or Impact	A072201	High Impedance		
7586	2022	All other reportable incidents	35039	Epicardial pacing lead	CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070101	Failure to Capture	06/12/2021	06/12/2021
									A070909	Oversensing	06/12/2021	06/12/2021
									A072201	High Impedance	06/12/2021	06/12/2021
									A072202	Low Impedance	06/12/2021	06/12/2021
									A090801	Signal Artifact/Noise	06/12/2021	06/12/2021
							F08	Hospitalization or Prolonged Hospitalization	A070101	Failure to Capture	06/12/2021	06/12/2021
									A070909	Oversensing	06/12/2021	06/12/2021
									A072201	High Impedance	06/12/2021	06/12/2021
									A072202	Low Impedance	06/12/2021	06/12/2021
									A090801	Signal Artifact/Noise	06/12/2021	06/12/2021
							F1903	Device Explantation	A070101	Failure to Capture	06/12/2021	06/12/2021
									A070909	Oversensing	06/12/2021	06/12/2021
									A072201	High Impedance	06/12/2021	06/12/2021
									A072202	Low Impedance	06/12/2021	06/12/2021
									A090801	Signal Artifact/Noise	06/12/2021	06/12/2021
							F23	Unexpected Medical Intervention	A070101	Failure to Capture	06/12/2021	06/12/2021
									A070909	Oversensing	06/12/2021	06/12/2021
									A072201	High Impedance	06/12/2021	06/12/2021
									A072202	Low Impedance	06/12/2021	06/12/2021
									A090801	Signal Artifact/Noise	06/12/2021	06/12/2021
7587	2022	All other reportable incidents	35039	Epicardial pacing lead	CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070909	Oversensing	06/12/2021	06/12/2021
									A072201	High Impedance	06/12/2021	06/12/2021
							F08	Hospitalization or Prolonged Hospitalization	A070909	Oversensing	06/12/2021	06/12/2021
									A072201	High Impedance	06/12/2021	06/12/2021

							F12	Serious Injury/ Illness/ Impairment	A070909	Oversensing	06/12/2021	06/12/2021
									A072201	High Impedance	06/12/2021	06/12/2021
							F1905	Device Revision or Replacement	A070909	Oversensing	06/12/2021	06/12/2021
									A072201	High Impedance	06/12/2021	06/12/2021
							F26	No Health Consequences or Impact	A070909	Oversensing	06/12/2021	06/12/2021
									A072201	High Impedance	06/12/2021	06/12/2021
7588	2022	All other reportable incidents	47265	Dual-chamber implantable pacemaker, rate-responsive	CT141	Electrodes/Leads and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem	10/02/2022	10/02/2022
									A070504	Premature Discharge of Battery	10/02/2022	10/02/2022
							F08	Hospitalization or Prolonged Hospitalization	A0705	Battery Problem	10/02/2022	10/02/2022
									A070504	Premature Discharge of Battery	10/02/2022	10/02/2022
							F1903	Device Explantation	A0705	Battery Problem	10/02/2022	10/02/2022
									A070504	Premature Discharge of Battery	10/02/2022	10/02/2022
					CT145	Cardiovascular devices	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem	10/02/2022	10/02/2022
									A070504	Premature Discharge of Battery	10/02/2022	10/02/2022
							F08	Hospitalization or Prolonged Hospitalization	A0705	Battery Problem	10/02/2022	10/02/2022
									A070504	Premature Discharge of Battery	10/02/2022	10/02/2022
							F1903	Device Explantation	A0705	Battery Problem	10/02/2022	10/02/2022
									A070504	Premature Discharge of Battery	10/02/2022	10/02/2022
					CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem	10/02/2022	10/02/2022
									A070504	Premature Discharge of Battery	10/02/2022	10/02/2022
							F08	Hospitalization or Prolonged Hospitalization	A0705	Battery Problem	10/02/2022	10/02/2022
									A070504	Premature Discharge of Battery	10/02/2022	10/02/2022
							F1903	Device Explantation	A0705	Battery Problem	10/02/2022	10/02/2022
									A070504	Premature Discharge of Battery	10/02/2022	10/02/2022
					CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem	10/02/2022	10/02/2022
									A070504	Premature Discharge of Battery	10/02/2022	10/02/2022
							F08	Hospitalization or Prolonged Hospitalization	A0705	Battery Problem	10/02/2022	10/02/2022
									A070504	Premature Discharge of Battery	10/02/2022	10/02/2022
							F1903	Device Explantation	A0705	Battery Problem	10/02/2022	10/02/2022
									A070504	Premature Discharge of Battery	10/02/2022	10/02/2022
					CT233	Surgical	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem	10/02/2022	10/02/2022
									A070504	Premature Discharge of Battery	10/02/2022	10/02/2022
							F08	Hospitalization or Prolonged Hospitalization	A0705	Battery Problem	10/02/2022	10/02/2022
									A070504	Premature Discharge of Battery	10/02/2022	10/02/2022
							F1903	Device Explantation	A0705	Battery Problem	10/02/2022	10/02/2022
									A070504	Premature Discharge of Battery	10/02/2022	10/02/2022
					CT2406	Implantable	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem	10/02/2022	10/02/2022
									A070504	Premature Discharge of Battery	10/02/2022	10/02/2022
							F08	Hospitalization or Prolonged Hospitalization	A0705	Battery Problem	10/02/2022	10/02/2022
									A070504	Premature Discharge of Battery	10/02/2022	10/02/2022
							F1903	Device Explantation	A0705	Battery Problem	10/02/2022	10/02/2022
									A070504	Premature Discharge of Battery	10/02/2022	10/02/2022
					CT2407	Active implantable	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem	10/02/2022	10/02/2022
									A070504	Premature Discharge of Battery	10/02/2022	10/02/2022
							F08	Hospitalization or Prolonged Hospitalization	A0705	Battery Problem	10/02/2022	10/02/2022
									A070504	Premature Discharge of Battery	10/02/2022	10/02/2022
							F1903	Device Explantation	A0705	Battery Problem	10/02/2022	10/02/2022
									A070504	Premature Discharge of Battery	10/02/2022	10/02/2022
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem	10/02/2022	10/02/2022
									A070504	Premature Discharge of Battery	10/02/2022	10/02/2022
							F08	Hospitalization or Prolonged Hospitalization	A0705	Battery Problem	10/02/2022	10/02/2022
									A070504	Premature Discharge of Battery	10/02/2022	10/02/2022
							F1903	Device Explantation	A0705	Battery Problem	10/02/2022	10/02/2022
									A070504	Premature Discharge of Battery	10/02/2022	10/02/2022
					CT2727	Active implant devices/ancillaries	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem	10/02/2022	10/02/2022
									A070504	Premature Discharge of Battery	10/02/2022	10/02/2022
							F08	Hospitalization or Prolonged Hospitalization	A0705	Battery Problem	10/02/2022	10/02/2022
									A070504	Premature Discharge of Battery	10/02/2022	10/02/2022
							F1903	Device Explantation	A0705	Battery Problem	10/02/2022	10/02/2022
									A070504	Premature Discharge of Battery	10/02/2022	10/02/2022
					CT2750	Implantable cardiac electrical stimulators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem	10/02/2022	10/02/2022
									A070504	Premature Discharge of Battery	10/02/2022	10/02/2022
							F08	Hospitalization or Prolonged Hospitalization	A0705	Battery Problem	10/02/2022	10/02/2022
									A070504	Premature Discharge of Battery	10/02/2022	10/02/2022
							F1903	Device Explantation	A0705	Battery Problem	10/02/2022	10/02/2022
									A070504	Premature Discharge of Battery	10/02/2022	10/02/2022
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem	10/02/2022	10/02/2022
									A070504	Premature Discharge of Battery	10/02/2022	10/02/2022
							F08	Hospitalization or Prolonged Hospitalization	A0705	Battery Problem	10/02/2022	10/02/2022
									A070504	Premature Discharge of Battery	10/02/2022	10/02/2022
							F1903	Device Explantation	A0705	Battery Problem	10/02/2022	10/02/2022
									A070504	Premature Discharge of Battery	10/02/2022	10/02/2022
					CT2985	Active	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem	10/02/2022	10/02/2022
									A070504	Premature Discharge of Battery	10/02/2022	10/02/2022
							F08	Hospitalization or Prolonged Hospitalization	A0705	Battery Problem	10/02/2022	10/02/2022
									A070504	Premature Discharge of Battery	10/02/2022	10/02/2022
							F1903	Device Explantation	A0705	Battery Problem	10/02/2022	10/02/2022
									A070504	Premature Discharge of Battery	10/02/2022	10/02/2022
					CT3001	Implantable metal	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem	10/02/2022	10/02/2022
									A070504	Premature Discharge of Battery	10/02/2022	10/02/2022
							F08	Hospitalization or Prolonged Hospitalization	A0705	Battery Problem	10/02/2022	10/02/2022
									A070504	Premature Discharge of Battery	10/02/2022	10/02/2022
							F1903	Device Explantation	A0705	Battery Problem	10/02/2022	10/02/2022
									A070504	Premature Discharge of Battery	10/02/2022	10/02/2022
					CT3002	Implantable synthetic polymer	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem	10/02/2022	10/02/2022
									A070504	Premature Discharge of Battery	10/02/2022	10/02/2022
							F08	Hospitalization or Prolonged Hospitalization	A0705	Battery Problem	10/02/2022	10/02/2022
									A070504	Premature Discharge of Battery	10/02/2022	10/02/2022

							F1903	Device Explantation	A0705	Battery Problem	10/02/2022	10/02/2022
									A070504	Premature Discharge of Battery	10/02/2022	10/02/2022
					CT3006	Implantable inorganic material	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem	10/02/2022	10/02/2022
									A070504	Premature Discharge of Battery	10/02/2022	10/02/2022
							F08	Hospitalization or Prolonged Hospitalization	A0705	Battery Problem	10/02/2022	10/02/2022
									A070504	Premature Discharge of Battery	10/02/2022	10/02/2022
							F1903	Device Explantation	A0705	Battery Problem	10/02/2022	10/02/2022
									A070504	Premature Discharge of Battery	10/02/2022	10/02/2022
					CT314	Electrically powered	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem	10/02/2022	10/02/2022
									A070504	Premature Discharge of Battery	10/02/2022	10/02/2022
							F08	Hospitalization or Prolonged Hospitalization	A0705	Battery Problem	10/02/2022	10/02/2022
									A070504	Premature Discharge of Battery	10/02/2022	10/02/2022
							F1903	Device Explantation	A0705	Battery Problem	10/02/2022	10/02/2022
									A070504	Premature Discharge of Battery	10/02/2022	10/02/2022
					CT321	Long-term surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem	10/02/2022	10/02/2022
									A070504	Premature Discharge of Battery	10/02/2022	10/02/2022
							F08	Hospitalization or Prolonged Hospitalization	A0705	Battery Problem	10/02/2022	10/02/2022
									A070504	Premature Discharge of Battery	10/02/2022	10/02/2022
							F1903	Device Explantation	A0705	Battery Problem	10/02/2022	10/02/2022
									A070504	Premature Discharge of Battery	10/02/2022	10/02/2022
					CT449	Generators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem	10/02/2022	10/02/2022
									A070504	Premature Discharge of Battery	10/02/2022	10/02/2022
							F08	Hospitalization or Prolonged Hospitalization	A0705	Battery Problem	10/02/2022	10/02/2022
									A070504	Premature Discharge of Battery	10/02/2022	10/02/2022
							F1903	Device Explantation	A0705	Battery Problem	10/02/2022	10/02/2022
									A070504	Premature Discharge of Battery	10/02/2022	10/02/2022
					CT981	Single use	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem	10/02/2022	10/02/2022
									A070504	Premature Discharge of Battery	10/02/2022	10/02/2022
							F08	Hospitalization or Prolonged Hospitalization	A0705	Battery Problem	10/02/2022	10/02/2022
									A070504	Premature Discharge of Battery	10/02/2022	10/02/2022
							F1903	Device Explantation	A0705	Battery Problem	10/02/2022	10/02/2022
									A070504	Premature Discharge of Battery	10/02/2022	10/02/2022
					CT983	Surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem	10/02/2022	10/02/2022
									A070504	Premature Discharge of Battery	10/02/2022	10/02/2022
							F08	Hospitalization or Prolonged Hospitalization	A0705	Battery Problem	10/02/2022	10/02/2022
									A070504	Premature Discharge of Battery	10/02/2022	10/02/2022
							F1903	Device Explantation	A0705	Battery Problem	10/02/2022	10/02/2022
									A070504	Premature Discharge of Battery	10/02/2022	10/02/2022
7589	2022	All other reportable incidents	47265	Dual-chamber implantable pacemaker, rate-responsive	CT141	Electrodes/Leads and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem		
							F26	No Health Consequences or Impact	A0705	Battery Problem		
					CT145	Cardiovascular devices	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem		
							F26	No Health Consequences or Impact	A0705	Battery Problem		
					CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem		
							F26	No Health Consequences or Impact	A0705	Battery Problem		
					CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem		
							F26	No Health Consequences or Impact	A0705	Battery Problem		
					CT233	Surgical	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem		
							F26	No Health Consequences or Impact	A0705	Battery Problem		
					CT2406	Implantable	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem		
							F26	No Health Consequences or Impact	A0705	Battery Problem		
					CT2407	Active implantable	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem		
							F26	No Health Consequences or Impact	A0705	Battery Problem		
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem		
							F26	No Health Consequences or Impact	A0705	Battery Problem		
					CT2727	Active implant devices/ancillaries	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem		
							F26	No Health Consequences or Impact	A0705	Battery Problem		
					CT2750	Implantable cardiac electrical stimulators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem		
							F26	No Health Consequences or Impact	A0705	Battery Problem		
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem		
							F26	No Health Consequences or Impact	A0705	Battery Problem		
					CT2985	Active	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem		
							F26	No Health Consequences or Impact	A0705	Battery Problem		
					CT3001	Implantable metal	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem		
							F26	No Health Consequences or Impact	A0705	Battery Problem		
					CT3002	Implantable synthetic polymer	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem		
							F26	No Health Consequences or Impact	A0705	Battery Problem		
					CT3006	Implantable inorganic material	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem		
							F26	No Health Consequences or Impact	A0705	Battery Problem		
					CT314	Electrically powered	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem		
							F26	No Health Consequences or Impact	A0705	Battery Problem		
					CT321	Long-term surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem		
							F26	No Health Consequences or Impact	A0705	Battery Problem		
					CT449	Generators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem		
							F26	No Health Consequences or Impact	A0705	Battery Problem		
					CT981	Single use	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem		
							F26	No Health Consequences or Impact	A0705	Battery Problem		
					CT983	Surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem		
							F26	No Health Consequences or Impact	A0705	Battery Problem		
7590	2022	All other reportable incidents	47263	Cardiac resynchronization therapy implantable pacemaker	CT1684	Cardiac pacemakers and associated devices	E0112	Dizziness	A070909	Oversensing	20/08/2022	20/08/2022
									A0712	Pacing Problem	20/08/2022	20/08/2022
									A0908	Incorrect, Inadequate or Imprecise Result or Readings	20/08/2022	20/08/2022
									A160503	Inappropriate or Unexpected Reset	20/08/2022	20/08/2022
							E060101	Asystole	A070909	Oversensing	20/08/2022	20/08/2022
									A0712	Pacing Problem	20/08/2022	20/08/2022
									A0908	Incorrect, Inadequate or Imprecise Result or Readings	20/08/2022	20/08/2022
									A160503	Inappropriate or Unexpected Reset	20/08/2022	20/08/2022

							E2403	No Clinical Signs, Symptoms or Conditions	A070909	Oversensing	20/08/2022	20/08/2022
									A0712	Pacing Problem	20/08/2022	20/08/2022
									A0908	Incorrect, Inadequate or Imprecise Result or Readings	20/08/2022	20/08/2022
									A160503	Inappropriate or Unexpected Reset	20/08/2022	20/08/2022
							F08	Hospitalization or Prolonged Hospitalization	A070909	Oversensing	20/08/2022	20/08/2022
									A0712	Pacing Problem	20/08/2022	20/08/2022
									A0908	Incorrect, Inadequate or Imprecise Result or Readings	20/08/2022	20/08/2022
									A160503	Inappropriate or Unexpected Reset	20/08/2022	20/08/2022
							F1001	Absence of Treatment	A070909	Oversensing	20/08/2022	20/08/2022
									A0712	Pacing Problem	20/08/2022	20/08/2022
									A0908	Incorrect, Inadequate or Imprecise Result or Readings	20/08/2022	20/08/2022
									A160503	Inappropriate or Unexpected Reset	20/08/2022	20/08/2022
							F1905	Device Revision or Replacement	A070909	Oversensing	20/08/2022	20/08/2022
									A0712	Pacing Problem	20/08/2022	20/08/2022
									A0908	Incorrect, Inadequate or Imprecise Result or Readings	20/08/2022	20/08/2022
									A160503	Inappropriate or Unexpected Reset	20/08/2022	20/08/2022
						CT1685	Cardiac pacemakers	E0112	Dizziness	A070909	Oversensing	20/08/2022	20/08/2022
										A0712	Pacing Problem	20/08/2022	20/08/2022
										A0908	Incorrect, Inadequate or Imprecise Result or Readings	20/08/2022	20/08/2022
										A160503	Inappropriate or Unexpected Reset	20/08/2022	20/08/2022
								E060101	Asystole	A070909	Oversensing	20/08/2022	20/08/2022
										A0712	Pacing Problem	20/08/2022	20/08/2022
										A0908	Incorrect, Inadequate or Imprecise Result or Readings	20/08/2022	20/08/2022
										A160503	Inappropriate or Unexpected Reset	20/08/2022	20/08/2022
								E2403	No Clinical Signs, Symptoms or Conditions	A070909	Oversensing	20/08/2022	20/08/2022
										A0712	Pacing Problem	20/08/2022	20/08/2022
										A0908	Incorrect, Inadequate or Imprecise Result or Readings	20/08/2022	20/08/2022
										A160503	Inappropriate or Unexpected Reset	20/08/2022	20/08/2022
								F08	Hospitalization or Prolonged Hospitalization	A070909	Oversensing	20/08/2022	20/08/2022
										A0712	Pacing Problem	20/08/2022	20/08/2022
										A0908	Incorrect, Inadequate or Imprecise Result or Readings	20/08/2022	20/08/2022
										A160503	Inappropriate or Unexpected Reset	20/08/2022	20/08/2022
								F1001	Absence of Treatment	A070909	Oversensing	20/08/2022	20/08/2022
										A0712	Pacing Problem	20/08/2022	20/08/2022
										A0908	Incorrect, Inadequate or Imprecise Result or Readings	20/08/2022	20/08/2022
										A160503	Inappropriate or Unexpected Reset	20/08/2022	20/08/2022
								F1905	Device Revision or Replacement	A070909	Oversensing	20/08/2022	20/08/2022
										A0712	Pacing Problem	20/08/2022	20/08/2022
										A0908	Incorrect, Inadequate or Imprecise Result or Readings	20/08/2022	20/08/2022
										A160503	Inappropriate or Unexpected Reset	20/08/2022	20/08/2022
						CT257	Implantable cardiac pacemakers	E0112	Dizziness	A070909	Oversensing	20/08/2022	20/08/2022
										A0712	Pacing Problem	20/08/2022	20/08/2022
										A0908	Incorrect, Inadequate or Imprecise Result or Readings	20/08/2022	20/08/2022
										A160503	Inappropriate or Unexpected Reset	20/08/2022	20/08/2022
								E060101	Asystole	A070909	Oversensing	20/08/2022	20/08/2022
										A0712	Pacing Problem	20/08/2022	20/08/2022
										A0908	Incorrect, Inadequate or Imprecise Result or Readings	20/08/2022	20/08/2022
										A160503	Inappropriate or Unexpected Reset	20/08/2022	20/08/2022
								E2403	No Clinical Signs, Symptoms or Conditions	A070909	Oversensing	20/08/2022	20/08/2022
										A0712	Pacing Problem	20/08/2022	20/08/2022
										A0908	Incorrect, Inadequate or Imprecise Result or Readings	20/08/2022	20/08/2022
										A160503	Inappropriate or Unexpected Reset	20/08/2022	20/08/2022
								F08	Hospitalization or Prolonged Hospitalization	A070909	Oversensing	20/08/2022	20/08/2022
										A0712	Pacing Problem	20/08/2022	20/08/2022
										A0908	Incorrect, Inadequate or Imprecise Result or Readings	20/08/2022	20/08/2022
										A160503	Inappropriate or Unexpected Reset	20/08/2022	20/08/2022
								F1001	Absence of Treatment	A070909	Oversensing	20/08/2022	20/08/2022
										A0712	Pacing Problem	20/08/2022	20/08/2022
										A0908	Incorrect, Inadequate or Imprecise Result or Readings	20/08/2022	20/08/2022
										A160503	Inappropriate or Unexpected Reset	20/08/2022	20/08/2022
								F1905	Device Revision or Replacement	A070909	Oversensing	20/08/2022	20/08/2022
										A0712	Pacing Problem	20/08/2022	20/08/2022
										A0908	Incorrect, Inadequate or Imprecise Result or Readings	20/08/2022	20/08/2022
										A160503	Inappropriate or Unexpected Reset	20/08/2022	20/08/2022
						CT2866	Implantable cardiac electrical stimulators	E0112	Dizziness	A070909	Oversensing	20/08/2022	20/08/2022
										A0712	Pacing Problem	20/08/2022	20/08/2022
										A0908	Incorrect, Inadequate or Imprecise Result or Readings	20/08/2022	20/08/2022
										A160503	Inappropriate or Unexpected Reset	20/08/2022	20/08/2022
								E060101	Asystole	A070909	Oversensing	20/08/2022	20/08/2022
										A0712	Pacing Problem	20/08/2022	20/08/2022
										A0908	Incorrect, Inadequate or Imprecise Result or Readings	20/08/2022	20/08/2022
										A160503	Inappropriate or Unexpected Reset	20/08/2022	20/08/2022
								E2403	No Clinical Signs, Symptoms or Conditions	A070909	Oversensing	20/08/2022	20/08/2022
										A0712	Pacing Problem	20/08/2022	20/08/2022

									A0908	Incorrect, Inadequate or Imprecise Result or Readings	20/08/2022	20/08/2022
									A160503	Inappropriate or Unexpected Reset	20/08/2022	20/08/2022
							F08	Hospitalization or Prolonged Hospitalization	A070909	Oversensing	20/08/2022	20/08/2022
									A0712	Pacing Problem	20/08/2022	20/08/2022
									A0908	Incorrect, Inadequate or Imprecise Result or Readings	20/08/2022	20/08/2022
									A160503	Inappropriate or Unexpected Reset	20/08/2022	20/08/2022
							F1001	Absence of Treatment	A070909	Oversensing	20/08/2022	20/08/2022
									A0712	Pacing Problem	20/08/2022	20/08/2022
									A0908	Incorrect, Inadequate or Imprecise Result or Readings	20/08/2022	20/08/2022
									A160503	Inappropriate or Unexpected Reset	20/08/2022	20/08/2022
							F1905	Device Revision or Replacement	A070909	Oversensing	20/08/2022	20/08/2022
									A0712	Pacing Problem	20/08/2022	20/08/2022
									A0908	Incorrect, Inadequate or Imprecise Result or Readings	20/08/2022	20/08/2022
									A160503	Inappropriate or Unexpected Reset	20/08/2022	20/08/2022
7591	2022	All other reportable incidents	17882	Physiologic-monitoring defibrillation system	CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070803	Failure to Power Up		
							F27	Problem Identified During Non-Clinical Procedure	A070803	Failure to Power Up		
7592	2022	All other reportable incidents	47265	Dual-chamber implantable pacemaker, rate-responsive	CT141	Electrodes/Leads and associated devices	E0717	Dyspnea	A071102	Failure to Interrogate	26/08/2022	26/08/2022
							E2311	Discomfort	A071102	Failure to Interrogate	26/08/2022	26/08/2022
							E2312	Fatigue	A071102	Failure to Interrogate	26/08/2022	26/08/2022
							E2403	No Clinical Signs, Symptoms or Conditions	A071102	Failure to Interrogate	26/08/2022	26/08/2022
							F1901	Additional Surgery	A071102	Failure to Interrogate	26/08/2022	26/08/2022
					CT145	Cardiovascular devices	E0717	Dyspnea	A071102	Failure to Interrogate	26/08/2022	26/08/2022
							E2311	Discomfort	A071102	Failure to Interrogate	26/08/2022	26/08/2022
							E2312	Fatigue	A071102	Failure to Interrogate	26/08/2022	26/08/2022
							E2403	No Clinical Signs, Symptoms or Conditions	A071102	Failure to Interrogate	26/08/2022	26/08/2022
							F1901	Additional Surgery	A071102	Failure to Interrogate	26/08/2022	26/08/2022
					CT1684	Cardiac pacemakers and associated devices	E0717	Dyspnea	A071102	Failure to Interrogate	26/08/2022	26/08/2022
							E2311	Discomfort	A071102	Failure to Interrogate	26/08/2022	26/08/2022
							E2312	Fatigue	A071102	Failure to Interrogate	26/08/2022	26/08/2022
							E2403	No Clinical Signs, Symptoms or Conditions	A071102	Failure to Interrogate	26/08/2022	26/08/2022
							F1901	Additional Surgery	A071102	Failure to Interrogate	26/08/2022	26/08/2022
					CT1685	Cardiac pacemakers	E0717	Dyspnea	A071102	Failure to Interrogate	26/08/2022	26/08/2022
							E2311	Discomfort	A071102	Failure to Interrogate	26/08/2022	26/08/2022
							E2312	Fatigue	A071102	Failure to Interrogate	26/08/2022	26/08/2022
							E2403	No Clinical Signs, Symptoms or Conditions	A071102	Failure to Interrogate	26/08/2022	26/08/2022
							F1901	Additional Surgery	A071102	Failure to Interrogate	26/08/2022	26/08/2022
					CT233	Surgical	E0717	Dyspnea	A071102	Failure to Interrogate	26/08/2022	26/08/2022
							E2311	Discomfort	A071102	Failure to Interrogate	26/08/2022	26/08/2022
							E2312	Fatigue	A071102	Failure to Interrogate	26/08/2022	26/08/2022
							E2403	No Clinical Signs, Symptoms or Conditions	A071102	Failure to Interrogate	26/08/2022	26/08/2022
							F1901	Additional Surgery	A071102	Failure to Interrogate	26/08/2022	26/08/2022
					CT2406	Implantable	E0717	Dyspnea	A071102	Failure to Interrogate	26/08/2022	26/08/2022
							E2311	Discomfort	A071102	Failure to Interrogate	26/08/2022	26/08/2022
							E2312	Fatigue	A071102	Failure to Interrogate	26/08/2022	26/08/2022
							E2403	No Clinical Signs, Symptoms or Conditions	A071102	Failure to Interrogate	26/08/2022	26/08/2022
							F1901	Additional Surgery	A071102	Failure to Interrogate	26/08/2022	26/08/2022
					CT2407	Active implantable	E0717	Dyspnea	A071102	Failure to Interrogate	26/08/2022	26/08/2022
							E2311	Discomfort	A071102	Failure to Interrogate	26/08/2022	26/08/2022
							E2312	Fatigue	A071102	Failure to Interrogate	26/08/2022	26/08/2022
							E2403	No Clinical Signs, Symptoms or Conditions	A071102	Failure to Interrogate	26/08/2022	26/08/2022
							F1901	Additional Surgery	A071102	Failure to Interrogate	26/08/2022	26/08/2022
					CT257	Implantable cardiac pacemakers	E0717	Dyspnea	A071102	Failure to Interrogate	26/08/2022	26/08/2022
							E2311	Discomfort	A071102	Failure to Interrogate	26/08/2022	26/08/2022
							E2312	Fatigue	A071102	Failure to Interrogate	26/08/2022	26/08/2022
							E2403	No Clinical Signs, Symptoms or Conditions	A071102	Failure to Interrogate	26/08/2022	26/08/2022
							F1901	Additional Surgery	A071102	Failure to Interrogate	26/08/2022	26/08/2022
					CT2727	Active implant devices/ancillaries	E0717	Dyspnea	A071102	Failure to Interrogate	26/08/2022	26/08/2022
							E2311	Discomfort	A071102	Failure to Interrogate	26/08/2022	26/08/2022
							E2312	Fatigue	A071102	Failure to Interrogate	26/08/2022	26/08/2022
							E2403	No Clinical Signs, Symptoms or Conditions	A071102	Failure to Interrogate	26/08/2022	26/08/2022
							F1901	Additional Surgery	A071102	Failure to Interrogate	26/08/2022	26/08/2022
					CT2750	Implantable cardiac electrical stimulators and associated devices	E0717	Dyspnea	A071102	Failure to Interrogate	26/08/2022	26/08/2022
							E2311	Discomfort	A071102	Failure to Interrogate	26/08/2022	26/08/2022
							E2312	Fatigue	A071102	Failure to Interrogate	26/08/2022	26/08/2022
							E2403	No Clinical Signs, Symptoms or Conditions	A071102	Failure to Interrogate	26/08/2022	26/08/2022
							F1901	Additional Surgery	A071102	Failure to Interrogate	26/08/2022	26/08/2022
					CT2866	Implantable cardiac electrical stimulators	E0717	Dyspnea	A071102	Failure to Interrogate	26/08/2022	26/08/2022
							E2311	Discomfort	A071102	Failure to Interrogate	26/08/2022	26/08/2022
							E2312	Fatigue	A071102	Failure to Interrogate	26/08/2022	26/08/2022
							E2403	No Clinical Signs, Symptoms or Conditions	A071102	Failure to Interrogate	26/08/2022	26/08/2022
							F1901	Additional Surgery	A071102	Failure to Interrogate	26/08/2022	26/08/2022
					CT2985	Active	E0717	Dyspnea	A071102	Failure to Interrogate	26/08/2022	26/08/2022
							E2311	Discomfort	A071102	Failure to Interrogate	26/08/2022	26/08/2022
							E2312	Fatigue	A071102	Failure to Interrogate	26/08/2022	26/08/2022
							E2403	No Clinical Signs, Symptoms or Conditions	A071102	Failure to Interrogate	26/08/2022	26/08/2022
							F1901	Additional Surgery	A071102	Failure to Interrogate	26/08/2022	26/08/2022
					CT3001	Implantable metal	E0717	Dyspnea	A071102	Failure to Interrogate	26/08/2022	26/08/2022
							E2311	Discomfort	A071102	Failure to Interrogate	26/08/2022	26/08/2022
							E2312	Fatigue	A071102	Failure to Interrogate	26/08/2022	26/08/2022
							E2403	No Clinical Signs, Symptoms or Conditions	A071102	Failure to Interrogate	26/08/2022	26/08/2022
							F1901	Additional Surgery	A071102	Failure to Interrogate	26/08/2022	26/08/2022
					CT3002	Implantable synthetic polymer	E0717	Dyspnea	A071102	Failure to Interrogate	26/08/2022	26/08/2022
							E2311	Discomfort	A071102	Failure to Interrogate	26/08/2022	26/08/2022
							E2312	Fatigue	A071102	Failure to Interrogate	26/08/2022	26/08/2022
							E2403	No Clinical Signs, Symptoms or Conditions	A071102	Failure to Interrogate	26/08/2022	26/08/2022

							F1901	Additional Surgery	A071102	Failure to Interrogate	26/08/2022	26/08/2022
					CT3006	Implantable inorganic material	E0717	Dyspnea	A071102	Failure to Interrogate	26/08/2022	26/08/2022
							E2311	Discomfort	A071102	Failure to Interrogate	26/08/2022	26/08/2022
							E2312	Fatigue	A071102	Failure to Interrogate	26/08/2022	26/08/2022
							E2403	No Clinical Signs, Symptoms or Conditions	A071102	Failure to Interrogate	26/08/2022	26/08/2022
							F1901	Additional Surgery	A071102	Failure to Interrogate	26/08/2022	26/08/2022
					CT314	Electrically powered	E0717	Dyspnea	A071102	Failure to Interrogate	26/08/2022	26/08/2022
							E2311	Discomfort	A071102	Failure to Interrogate	26/08/2022	26/08/2022
							E2312	Fatigue	A071102	Failure to Interrogate	26/08/2022	26/08/2022
							E2403	No Clinical Signs, Symptoms or Conditions	A071102	Failure to Interrogate	26/08/2022	26/08/2022
							F1901	Additional Surgery	A071102	Failure to Interrogate	26/08/2022	26/08/2022
					CT321	Long-term surgical invasive	E0717	Dyspnea	A071102	Failure to Interrogate	26/08/2022	26/08/2022
							E2311	Discomfort	A071102	Failure to Interrogate	26/08/2022	26/08/2022
							E2312	Fatigue	A071102	Failure to Interrogate	26/08/2022	26/08/2022
							E2403	No Clinical Signs, Symptoms or Conditions	A071102	Failure to Interrogate	26/08/2022	26/08/2022
							F1901	Additional Surgery	A071102	Failure to Interrogate	26/08/2022	26/08/2022
					CT449	Generators and associated devices	E0717	Dyspnea	A071102	Failure to Interrogate	26/08/2022	26/08/2022
							E2311	Discomfort	A071102	Failure to Interrogate	26/08/2022	26/08/2022
							E2312	Fatigue	A071102	Failure to Interrogate	26/08/2022	26/08/2022
							E2403	No Clinical Signs, Symptoms or Conditions	A071102	Failure to Interrogate	26/08/2022	26/08/2022
							F1901	Additional Surgery	A071102	Failure to Interrogate	26/08/2022	26/08/2022
					CT981	Single use	E0717	Dyspnea	A071102	Failure to Interrogate	26/08/2022	26/08/2022
							E2311	Discomfort	A071102	Failure to Interrogate	26/08/2022	26/08/2022
							E2312	Fatigue	A071102	Failure to Interrogate	26/08/2022	26/08/2022
							E2403	No Clinical Signs, Symptoms or Conditions	A071102	Failure to Interrogate	26/08/2022	26/08/2022
							F1901	Additional Surgery	A071102	Failure to Interrogate	26/08/2022	26/08/2022
					CT983	Surgical invasive	E0717	Dyspnea	A071102	Failure to Interrogate	26/08/2022	26/08/2022
							E2311	Discomfort	A071102	Failure to Interrogate	26/08/2022	26/08/2022
							E2312	Fatigue	A071102	Failure to Interrogate	26/08/2022	26/08/2022
							E2403	No Clinical Signs, Symptoms or Conditions	A071102	Failure to Interrogate	26/08/2022	26/08/2022
							F1901	Additional Surgery	A071102	Failure to Interrogate	26/08/2022	26/08/2022
7593	2022	All other reportable incidents	47270	Cardiac resynchronization therapy implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A070101	Failure to Capture		
									A070909	Oversensing		
							F26	No Health Consequences or Impact	A070101	Failure to Capture		
									A070909	Oversensing		
7594	2022	All other reportable incidents	36241	Pacemaker/defibrillator lead	CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A051201	Device Dislodged or Dislocated		
							F2203	Imaging Required	A051201	Device Dislodged or Dislocated		
							F26	No Health Consequences or Impact	A051201	Device Dislodged or Dislocated		
7595	2022	All other reportable incidents	17882	Physiologic-monitoring defibrillation system	CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A20	Installation-Related Problem		
							F27	Problem Identified During Non-Clinical Procedure	A20	Installation-Related Problem		
7596	2022	All other reportable incidents	35223	Endocardial/interventricular septal pacing lead	CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070909	Oversensing	03/12/2021	03/12/2021
									A0712	Pacing Problem	03/12/2021	03/12/2021
									A090801	Signal Artifact/Noise	03/12/2021	03/12/2021
							F26	No Health Consequences or Impact	A070909	Oversensing	03/12/2021	03/12/2021
									A0712	Pacing Problem	03/12/2021	03/12/2021
									A090801	Signal Artifact/Noise	03/12/2021	03/12/2021
7597	2022	All other reportable incidents	47263	Cardiac resynchronization therapy implantable pacemaker	CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A0908	Incorrect, Inadequate or Imprecise Result or Readings	25/08/2022	25/08/2022
									A160503	Inappropriate or Unexpected Reset	25/08/2022	25/08/2022
							F08	Hospitalization or Prolonged Hospitalization	A0908	Incorrect, Inadequate or Imprecise Result or Readings	25/08/2022	25/08/2022
									A160503	Inappropriate or Unexpected Reset	25/08/2022	25/08/2022
							F1905	Device Revision or Replacement	A0908	Incorrect, Inadequate or Imprecise Result or Readings	25/08/2022	25/08/2022
									A160503	Inappropriate or Unexpected Reset	25/08/2022	25/08/2022
					CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A0908	Incorrect, Inadequate or Imprecise Result or Readings	25/08/2022	25/08/2022
									A160503	Inappropriate or Unexpected Reset	25/08/2022	25/08/2022
							F08	Hospitalization or Prolonged Hospitalization	A0908	Incorrect, Inadequate or Imprecise Result or Readings	25/08/2022	25/08/2022
									A160503	Inappropriate or Unexpected Reset	25/08/2022	25/08/2022
							F1905	Device Revision or Replacement	A0908	Incorrect, Inadequate or Imprecise Result or Readings	25/08/2022	25/08/2022
									A160503	Inappropriate or Unexpected Reset	25/08/2022	25/08/2022
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A0908	Incorrect, Inadequate or Imprecise Result or Readings	25/08/2022	25/08/2022
									A160503	Inappropriate or Unexpected Reset	25/08/2022	25/08/2022
							F08	Hospitalization or Prolonged Hospitalization	A0908	Incorrect, Inadequate or Imprecise Result or Readings	25/08/2022	25/08/2022
									A160503	Inappropriate or Unexpected Reset	25/08/2022	25/08/2022
							F1905	Device Revision or Replacement	A0908	Incorrect, Inadequate or Imprecise Result or Readings	25/08/2022	25/08/2022
									A160503	Inappropriate or Unexpected Reset	25/08/2022	25/08/2022
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A0908	Incorrect, Inadequate or Imprecise Result or Readings	25/08/2022	25/08/2022
									A160503	Inappropriate or Unexpected Reset	25/08/2022	25/08/2022
							F08	Hospitalization or Prolonged Hospitalization	A0908	Incorrect, Inadequate or Imprecise Result or Readings	25/08/2022	25/08/2022
									A160503	Inappropriate or Unexpected Reset	25/08/2022	25/08/2022
							F1905	Device Revision or Replacement	A0908	Incorrect, Inadequate or Imprecise Result or Readings	25/08/2022	25/08/2022
									A160503	Inappropriate or Unexpected Reset	25/08/2022	25/08/2022
7598	2022	All other reportable incidents	35852	Single-chamber implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	06/12/2022	06/12/2022
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	06/12/2022	06/12/2022
							F1903	Device Explantation	A070504	Premature Discharge of Battery	06/12/2022	06/12/2022
7599	2022	All other reportable incidents	47265	Dual-chamber implantable pacemaker, rate-responsive	CT141	Electrodes/Leads and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	03/12/2021	03/12/2021
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	03/12/2021	03/12/2021
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	03/12/2021	03/12/2021
					CT145	Cardiovascular devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	03/12/2021	03/12/2021
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	03/12/2021	03/12/2021
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	03/12/2021	03/12/2021

					CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	03/12/2021	03/12/2021
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	03/12/2021	03/12/2021
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	03/12/2021	03/12/2021
					CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	03/12/2021	03/12/2021
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	03/12/2021	03/12/2021
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	03/12/2021	03/12/2021
					CT233	Surgical	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	03/12/2021	03/12/2021
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	03/12/2021	03/12/2021
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	03/12/2021	03/12/2021
					CT2406	Implantable	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	03/12/2021	03/12/2021
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	03/12/2021	03/12/2021
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	03/12/2021	03/12/2021
					CT2407	Active implantable	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	03/12/2021	03/12/2021
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	03/12/2021	03/12/2021
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	03/12/2021	03/12/2021
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	03/12/2021	03/12/2021
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	03/12/2021	03/12/2021
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	03/12/2021	03/12/2021
					CT2727	Active implant devices/ancillaries	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	03/12/2021	03/12/2021
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	03/12/2021	03/12/2021
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	03/12/2021	03/12/2021
					CT2750	Implantable cardiac electrical stimulators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	03/12/2021	03/12/2021
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	03/12/2021	03/12/2021
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	03/12/2021	03/12/2021
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	03/12/2021	03/12/2021
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	03/12/2021	03/12/2021
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	03/12/2021	03/12/2021
					CT2985	Active	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	03/12/2021	03/12/2021
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	03/12/2021	03/12/2021
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	03/12/2021	03/12/2021
					CT3001	Implantable metal	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	03/12/2021	03/12/2021
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	03/12/2021	03/12/2021
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	03/12/2021	03/12/2021
					CT3002	Implantable synthetic polymer	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	03/12/2021	03/12/2021
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	03/12/2021	03/12/2021
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	03/12/2021	03/12/2021
					CT3006	Implantable inorganic material	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	03/12/2021	03/12/2021
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	03/12/2021	03/12/2021
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	03/12/2021	03/12/2021
					CT314	Electrically powered	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	03/12/2021	03/12/2021
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	03/12/2021	03/12/2021
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	03/12/2021	03/12/2021
					CT321	Long-term surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	03/12/2021	03/12/2021
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	03/12/2021	03/12/2021
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	03/12/2021	03/12/2021
					CT449	Generators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	03/12/2021	03/12/2021
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	03/12/2021	03/12/2021
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	03/12/2021	03/12/2021
					CT981	Single use	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	03/12/2021	03/12/2021
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	03/12/2021	03/12/2021
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	03/12/2021	03/12/2021
					CT983	Surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	03/12/2021	03/12/2021
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	03/12/2021	03/12/2021
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	03/12/2021	03/12/2021
7600	2022	All other reportable incidents	35223	Endocardial/interventricular septal pacing lead	CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A090801	Signal Artifact/Noise		
							F26	No Health Consequences or Impact	A090801	Signal Artifact/Noise		
7601	2022	All other reportable incidents	35852	Single-chamber implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem	28/12/2022	28/12/2022
									A070504	Premature Discharge of Battery	28/12/2022	28/12/2022
							F08	Hospitalization or Prolonged Hospitalization	A0705	Battery Problem	28/12/2022	28/12/2022
									A070504	Premature Discharge of Battery	28/12/2022	28/12/2022
							F1903	Device Explantation	A0705	Battery Problem	28/12/2022	28/12/2022
									A070504	Premature Discharge of Battery	28/12/2022	28/12/2022
7602	2022	All other reportable incidents	17882	Physiologic-monitoring defibrillation system	CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070908	Failure to Sense		
							F26	No Health Consequences or Impact	A070908	Failure to Sense		
7603	2022	All other reportable incidents	17882	Physiologic-monitoring defibrillation system	CT1684	Cardiac pacemakers and associated devices	E0602	Cardiac Arrest	A12	Connection Problem		
							E2403	No Clinical Signs, Symptoms or Conditions	A12	Connection Problem		
							F26	No Health Consequences or Impact	A12	Connection Problem		
7604	2022	All other reportable incidents	17882	Physiologic-monitoring defibrillation system	CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070908	Failure to Sense		
							F26	No Health Consequences or Impact	A070908	Failure to Sense		
7605	2022	All other reportable incidents	47270	Cardiac resynchronization therapy implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E2104	Electric Shock	A070909	Oversensing		
									A0712	Pacing Problem		
									A071302	Inappropriate/Inadequate Shock/Stimulation		
									A090801	Signal Artifact/Noise		
							E2330	Pain	A070909	Oversensing		
									A0712	Pacing Problem		
									A071302	Inappropriate/Inadequate Shock/Stimulation		
									A090801	Signal Artifact/Noise		
							E2403	No Clinical Signs, Symptoms or Conditions	A070909	Oversensing		
									A0712	Pacing Problem		
									A071302	Inappropriate/Inadequate Shock/Stimulation		
									A090801	Signal Artifact/Noise		
							F08	Hospitalization or Prolonged Hospitalization	A070909	Oversensing		
									A0712	Pacing Problem		
									A071302	Inappropriate/Inadequate Shock/Stimulation		
									A090801	Signal Artifact/Noise		
							F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A070909	Oversensing		
									A0712	Pacing Problem		
									A071302	Inappropriate/Inadequate Shock/Stimulation		

									A090801	Signal Artifact/Noise		
7606	2022	All other reportable incidents	36241	Pacemaker/defibrillator lead	CT1684	Cardiac pacemakers and associated devices	E2104	Electric Shock	A070909	Oversensing		
									A0712	Pacing Problem		
									A071302	Inappropriate/Inadequate Shock/Stimulation		
									A090801	Signal Artifact/Noise		
							E2330	Pain	A070909	Oversensing		
									A0712	Pacing Problem		
									A071302	Inappropriate/Inadequate Shock/Stimulation		
									A090801	Signal Artifact/Noise		
							E2403	No Clinical Signs, Symptoms or Conditions	A070909	Oversensing		
									A0712	Pacing Problem		
									A071302	Inappropriate/Inadequate Shock/Stimulation		
									A090801	Signal Artifact/Noise		
							F08	Hospitalization or Prolonged Hospitalization	A070909	Oversensing		
									A0712	Pacing Problem		
									A071302	Inappropriate/Inadequate Shock/Stimulation		
									A090801	Signal Artifact/Noise		
							F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A070909	Oversensing		
									A0712	Pacing Problem		
									A071302	Inappropriate/Inadequate Shock/Stimulation		
									A090801	Signal Artifact/Noise		
7607	2022	Unanticipated serious deterioration in state of health	36241	Pacemaker/defibrillator lead	CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A0203	Defective Device		
									A072202	Low Impedance		
							F08	Hospitalization or Prolonged Hospitalization	A0203	Defective Device		
									A072202	Low Impedance		
							F1905	Device Revision or Replacement	A0203	Defective Device		
									A072202	Low Impedance		
7608	2022	All other reportable incidents	36241	Pacemaker/defibrillator lead	CT1684	Cardiac pacemakers and associated devices	E0119	Loss of Consciousness	A24	No Apparent Device Problem	09/12/2021	09/12/2021
							E0604	Cardiac Perforation	A24	No Apparent Device Problem	09/12/2021	09/12/2021
							F08	Hospitalization or Prolonged Hospitalization	A24	No Apparent Device Problem	09/12/2021	09/12/2021
							F1905	Device Revision or Replacement	A24	No Apparent Device Problem	09/12/2021	09/12/2021
7609	2022	All other reportable incidents	37265	Dual-chamber implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A072201	High Impedance		
							F26	No Health Consequences or Impact	A072201	High Impedance		
7610	2022	All other reportable incidents	36241	Pacemaker/defibrillator lead	CT1684	Cardiac pacemakers and associated devices	E0604	Cardiac Perforation	A24	No Apparent Device Problem		
							E0619	Pericardial Effusion	A24	No Apparent Device Problem		
							E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem		
							F08	Hospitalization or Prolonged Hospitalization	A24	No Apparent Device Problem		
							F1904	Device Repositioning	A24	No Apparent Device Problem		
7611	2022	All other reportable incidents	36241	Pacemaker/defibrillator lead	CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070102	High Capture Threshold		
									A072201	High Impedance		
							F26	No Health Consequences or Impact	A070102	High Capture Threshold		
									A072201	High Impedance		
7612	2022	All other reportable incidents	35852	Single-chamber implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
									A1301	Failure to Read Input Signal		
									A23	Use of Device Problem		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
									A1301	Failure to Read Input Signal		
									A23	Use of Device Problem		
7613	2022	All other reportable incidents	60789	Intracardiac pacemaker	CT130	Catheters and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070102	High Capture Threshold	30/08/2022	30/08/2022
							F08	Hospitalization or Prolonged Hospitalization	A070102	High Capture Threshold	30/08/2022	30/08/2022
							F12	Serious Injury/ Illness/ Impairment	A070102	High Capture Threshold	30/08/2022	30/08/2022
							F1905	Device Revision or Replacement	A070102	High Capture Threshold	30/08/2022	30/08/2022
					CT145	Cardiovascular devices	E2403	No Clinical Signs, Symptoms or Conditions	A070102	High Capture Threshold	30/08/2022	30/08/2022
							F08	Hospitalization or Prolonged Hospitalization	A070102	High Capture Threshold	30/08/2022	30/08/2022
							F12	Serious Injury/ Illness/ Impairment	A070102	High Capture Threshold	30/08/2022	30/08/2022
							F1905	Device Revision or Replacement	A070102	High Capture Threshold	30/08/2022	30/08/2022
					CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070102	High Capture Threshold	30/08/2022	30/08/2022
							F08	Hospitalization or Prolonged Hospitalization	A070102	High Capture Threshold	30/08/2022	30/08/2022
							F12	Serious Injury/ Illness/ Impairment	A070102	High Capture Threshold	30/08/2022	30/08/2022
							F1905	Device Revision or Replacement	A070102	High Capture Threshold	30/08/2022	30/08/2022
					CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A070102	High Capture Threshold	30/08/2022	30/08/2022
							F08	Hospitalization or Prolonged Hospitalization	A070102	High Capture Threshold	30/08/2022	30/08/2022
							F12	Serious Injury/ Illness/ Impairment	A070102	High Capture Threshold	30/08/2022	30/08/2022
							F1905	Device Revision or Replacement	A070102	High Capture Threshold	30/08/2022	30/08/2022
					CT233	Surgical	E2403	No Clinical Signs, Symptoms or Conditions	A070102	High Capture Threshold	30/08/2022	30/08/2022
							F08	Hospitalization or Prolonged Hospitalization	A070102	High Capture Threshold	30/08/2022	30/08/2022
							F12	Serious Injury/ Illness/ Impairment	A070102	High Capture Threshold	30/08/2022	30/08/2022
							F1905	Device Revision or Replacement	A070102	High Capture Threshold	30/08/2022	30/08/2022
					CT2406	Implantable	E2403	No Clinical Signs, Symptoms or Conditions	A070102	High Capture Threshold	30/08/2022	30/08/2022
							F08	Hospitalization or Prolonged Hospitalization	A070102	High Capture Threshold	30/08/2022	30/08/2022
							F12	Serious Injury/ Illness/ Impairment	A070102	High Capture Threshold	30/08/2022	30/08/2022
							F1905	Device Revision or Replacement	A070102	High Capture Threshold	30/08/2022	30/08/2022
					CT2407	Active implantable	E2403	No Clinical Signs, Symptoms or Conditions	A070102	High Capture Threshold	30/08/2022	30/08/2022
							F08	Hospitalization or Prolonged Hospitalization	A070102	High Capture Threshold	30/08/2022	30/08/2022
							F12	Serious Injury/ Illness/ Impairment	A070102	High Capture Threshold	30/08/2022	30/08/2022
							F1905	Device Revision or Replacement	A070102	High Capture Threshold	30/08/2022	30/08/2022
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A070102	High Capture Threshold	30/08/2022	30/08/2022
							F08	Hospitalization or Prolonged Hospitalization	A070102	High Capture Threshold	30/08/2022	30/08/2022
							F12	Serious Injury/ Illness/ Impairment	A070102	High Capture Threshold	30/08/2022	30/08/2022
							F1905	Device Revision or Replacement	A070102	High Capture Threshold	30/08/2022	30/08/2022
					CT2727	Active implant devices/ancillaries	E2403	No Clinical Signs, Symptoms or Conditions	A070102	High Capture Threshold	30/08/2022	30/08/2022
							F08	Hospitalization or Prolonged Hospitalization	A070102	High Capture Threshold	30/08/2022	30/08/2022
							F12	Serious Injury/ Illness/ Impairment	A070102	High Capture Threshold	30/08/2022	30/08/2022
							F1905	Device Revision or Replacement	A070102	High Capture Threshold	30/08/2022	30/08/2022
					CT2750	Implantable cardiac electrical stimulators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070102	High Capture Threshold	30/08/2022	30/08/2022
							F08	Hospitalization or Prolonged Hospitalization	A070102	High Capture Threshold	30/08/2022	30/08/2022
							F12	Serious Injury/ Illness/ Impairment	A070102	High Capture Threshold	30/08/2022	30/08/2022
							F1905	Device Revision or Replacement	A070102	High Capture Threshold	30/08/2022	30/08/2022

					CT2818	Catheterization/Cannulation/Fluid path devices	E2403	No Clinical Signs, Symptoms or Conditions	A070102	High Capture Threshold	30/08/2022	30/08/2022
							F08	Hospitalization or Prolonged Hospitalization	A070102	High Capture Threshold	30/08/2022	30/08/2022
							F12	Serious Injury/ Illness/ Impairment	A070102	High Capture Threshold	30/08/2022	30/08/2022
							F1905	Device Revision or Replacement	A070102	High Capture Threshold	30/08/2022	30/08/2022
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A070102	High Capture Threshold	30/08/2022	30/08/2022
							F08	Hospitalization or Prolonged Hospitalization	A070102	High Capture Threshold	30/08/2022	30/08/2022
							F12	Serious Injury/ Illness/ Impairment	A070102	High Capture Threshold	30/08/2022	30/08/2022
							F1905	Device Revision or Replacement	A070102	High Capture Threshold	30/08/2022	30/08/2022
					CT2985	Active	E2403	No Clinical Signs, Symptoms or Conditions	A070102	High Capture Threshold	30/08/2022	30/08/2022
							F08	Hospitalization or Prolonged Hospitalization	A070102	High Capture Threshold	30/08/2022	30/08/2022
							F12	Serious Injury/ Illness/ Impairment	A070102	High Capture Threshold	30/08/2022	30/08/2022
							F1905	Device Revision or Replacement	A070102	High Capture Threshold	30/08/2022	30/08/2022
					CT3001	Implantable metal	E2403	No Clinical Signs, Symptoms or Conditions	A070102	High Capture Threshold	30/08/2022	30/08/2022
							F08	Hospitalization or Prolonged Hospitalization	A070102	High Capture Threshold	30/08/2022	30/08/2022
							F12	Serious Injury/ Illness/ Impairment	A070102	High Capture Threshold	30/08/2022	30/08/2022
							F1905	Device Revision or Replacement	A070102	High Capture Threshold	30/08/2022	30/08/2022
					CT3002	Implantable synthetic polymer	E2403	No Clinical Signs, Symptoms or Conditions	A070102	High Capture Threshold	30/08/2022	30/08/2022
							F08	Hospitalization or Prolonged Hospitalization	A070102	High Capture Threshold	30/08/2022	30/08/2022
							F12	Serious Injury/ Illness/ Impairment	A070102	High Capture Threshold	30/08/2022	30/08/2022
							F1905	Device Revision or Replacement	A070102	High Capture Threshold	30/08/2022	30/08/2022
					CT3006	Implantable inorganic material	E2403	No Clinical Signs, Symptoms or Conditions	A070102	High Capture Threshold	30/08/2022	30/08/2022
							F08	Hospitalization or Prolonged Hospitalization	A070102	High Capture Threshold	30/08/2022	30/08/2022
							F12	Serious Injury/ Illness/ Impairment	A070102	High Capture Threshold	30/08/2022	30/08/2022
							F1905	Device Revision or Replacement	A070102	High Capture Threshold	30/08/2022	30/08/2022
					CT314	Electrically powered	E2403	No Clinical Signs, Symptoms or Conditions	A070102	High Capture Threshold	30/08/2022	30/08/2022
							F08	Hospitalization or Prolonged Hospitalization	A070102	High Capture Threshold	30/08/2022	30/08/2022
							F12	Serious Injury/ Illness/ Impairment	A070102	High Capture Threshold	30/08/2022	30/08/2022
							F1905	Device Revision or Replacement	A070102	High Capture Threshold	30/08/2022	30/08/2022
					CT321	Long-term surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A070102	High Capture Threshold	30/08/2022	30/08/2022
							F08	Hospitalization or Prolonged Hospitalization	A070102	High Capture Threshold	30/08/2022	30/08/2022
							F12	Serious Injury/ Illness/ Impairment	A070102	High Capture Threshold	30/08/2022	30/08/2022
							F1905	Device Revision or Replacement	A070102	High Capture Threshold	30/08/2022	30/08/2022
					CT981	Single use	E2403	No Clinical Signs, Symptoms or Conditions	A070102	High Capture Threshold	30/08/2022	30/08/2022
							F08	Hospitalization or Prolonged Hospitalization	A070102	High Capture Threshold	30/08/2022	30/08/2022
							F12	Serious Injury/ Illness/ Impairment	A070102	High Capture Threshold	30/08/2022	30/08/2022
							F1905	Device Revision or Replacement	A070102	High Capture Threshold	30/08/2022	30/08/2022
					CT983	Surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A070102	High Capture Threshold	30/08/2022	30/08/2022
							F08	Hospitalization or Prolonged Hospitalization	A070102	High Capture Threshold	30/08/2022	30/08/2022
							F12	Serious Injury/ Illness/ Impairment	A070102	High Capture Threshold	30/08/2022	30/08/2022
							F1905	Device Revision or Replacement	A070102	High Capture Threshold	30/08/2022	30/08/2022
7614	2022	All other reportable incidents	47265	Dual-chamber implantable pacemaker, rate-responsive	CT141	Electrodes/Leads and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A090805	Incorrect Measurement	14/12/2021	14/12/2021
							F1901	Additional Surgery	A090805	Incorrect Measurement	14/12/2021	14/12/2021
					CT145	Cardiovascular devices	E2403	No Clinical Signs, Symptoms or Conditions	A090805	Incorrect Measurement	14/12/2021	14/12/2021
							F1901	Additional Surgery	A090805	Incorrect Measurement	14/12/2021	14/12/2021
					CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A090805	Incorrect Measurement	14/12/2021	14/12/2021
							F1901	Additional Surgery	A090805	Incorrect Measurement	14/12/2021	14/12/2021
					CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A090805	Incorrect Measurement	14/12/2021	14/12/2021
							F1901	Additional Surgery	A090805	Incorrect Measurement	14/12/2021	14/12/2021
					CT233	Surgical	E2403	No Clinical Signs, Symptoms or Conditions	A090805	Incorrect Measurement	14/12/2021	14/12/2021
							F1901	Additional Surgery	A090805	Incorrect Measurement	14/12/2021	14/12/2021
					CT2406	Implantable	E2403	No Clinical Signs, Symptoms or Conditions	A090805	Incorrect Measurement	14/12/2021	14/12/2021
							F1901	Additional Surgery	A090805	Incorrect Measurement	14/12/2021	14/12/2021
					CT2407	Active implantable	E2403	No Clinical Signs, Symptoms or Conditions	A090805	Incorrect Measurement	14/12/2021	14/12/2021
							F1901	Additional Surgery	A090805	Incorrect Measurement	14/12/2021	14/12/2021
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A090805	Incorrect Measurement	14/12/2021	14/12/2021
							F1901	Additional Surgery	A090805	Incorrect Measurement	14/12/2021	14/12/2021
					CT2727	Active implant devices/ancillaries	E2403	No Clinical Signs, Symptoms or Conditions	A090805	Incorrect Measurement	14/12/2021	14/12/2021
							F1901	Additional Surgery	A090805	Incorrect Measurement	14/12/2021	14/12/2021
					CT2750	Implantable cardiac electrical stimulators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A090805	Incorrect Measurement	14/12/2021	14/12/2021
							F1901	Additional Surgery	A090805	Incorrect Measurement	14/12/2021	14/12/2021
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A090805	Incorrect Measurement	14/12/2021	14/12/2021
							F1901	Additional Surgery	A090805	Incorrect Measurement	14/12/2021	14/12/2021
					CT2985	Active	E2403	No Clinical Signs, Symptoms or Conditions	A090805	Incorrect Measurement	14/12/2021	14/12/2021
							F1901	Additional Surgery	A090805	Incorrect Measurement	14/12/2021	14/12/2021
					CT3001	Implantable metal	E2403	No Clinical Signs, Symptoms or Conditions	A090805	Incorrect Measurement	14/12/2021	14/12/2021
							F1901	Additional Surgery	A090805	Incorrect Measurement	14/12/2021	14/12/2021
					CT3002	Implantable synthetic polymer	E2403	No Clinical Signs, Symptoms or Conditions	A090805	Incorrect Measurement	14/12/2021	14/12/2021
							F1901	Additional Surgery	A090805	Incorrect Measurement	14/12/2021	14/12/2021
					CT3006	Implantable inorganic material	E2403	No Clinical Signs, Symptoms or Conditions	A090805	Incorrect Measurement	14/12/2021	14/12/2021
							F1901	Additional Surgery	A090805	Incorrect Measurement	14/12/2021	14/12/2021
					CT314	Electrically powered	E2403	No Clinical Signs, Symptoms or Conditions	A090805	Incorrect Measurement	14/12/2021	14/12/2021
							F1901	Additional Surgery	A090805	Incorrect Measurement	14/12/2021	14/12/2021
					CT321	Long-term surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A090805	Incorrect Measurement	14/12/2021	14/12/2021
							F1901	Additional Surgery	A090805	Incorrect Measurement	14/12/2021	14/12/2021
					CT449	Generators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A090805	Incorrect Measurement	14/12/2021	14/12/2021
							F1901	Additional Surgery	A090805	Incorrect Measurement	14/12/2021	14/12/2021
					CT981	Single use	E2403	No Clinical Signs, Symptoms or Conditions	A090805	Incorrect Measurement	14/12/2021	14/12/2021
							F1901	Additional Surgery	A090805	Incorrect Measurement	14/12/2021	14/12/2021
					CT983	Surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A090805	Incorrect Measurement	14/12/2021	14/12/2021
							F1901	Additional Surgery	A090805	Incorrect Measurement	14/12/2021	14/12/2021
7615	2022	All other reportable incidents	35223	Endocardial/interventricular septal pacing lead	CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070909	Oversensing		
									A0722	Impedance Problem		
									A072201	High Impedance		
									A090801	Signal Artifact/Noise		
									A1205	Loose or Intermittent Connection		
							F26	No Health Consequences or Impact	A070909	Oversensing		
									A0722	Impedance Problem		
									A072201	High Impedance		

									A090801	Signal Artifact/Noise		
									A1205	Loose or Intermittent Connection		
7616	2022	All other reportable incidents	17882	Physiologic-monitoring defibrillation system	CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A1203	Disconnection		
							F27	Problem Identified During Non-Clinical Procedure	A1203	Disconnection		
7617	2022	Death	17882	Physiologic-monitoring defibrillation system	CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A071301	Failure to Deliver Shock/Stimulation		
							F02	Death	A071301	Failure to Deliver Shock/Stimulation		
7618	2022	All other reportable incidents	35852	Single-chamber implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E2104	Electric Shock	A071302	Inappropriate/Inadequate Shock/Stimulation		
									A0722	Impedance Problem		
									A072201	High Impedance		
							E2403	No Clinical Signs, Symptoms or Conditions	A071302	Inappropriate/Inadequate Shock/Stimulation		
									A0722	Impedance Problem		
									A072201	High Impedance		
							F08	Hospitalization or Prolonged Hospitalization	A071302	Inappropriate/Inadequate Shock/Stimulation		
									A0722	Impedance Problem		
									A072201	High Impedance		
							F2303	Medication Required	A071302	Inappropriate/Inadequate Shock/Stimulation		
									A0722	Impedance Problem		
									A072201	High Impedance		
7619	2022	All other reportable incidents	36241	Pacemaker/defibrillator lead	CT1684	Cardiac pacemakers and associated devices	E2104	Electric Shock	A071302	Inappropriate/Inadequate Shock/Stimulation		
									A0722	Impedance Problem		
									A072201	High Impedance		
							E2403	No Clinical Signs, Symptoms or Conditions	A071302	Inappropriate/Inadequate Shock/Stimulation		
									A0722	Impedance Problem		
									A072201	High Impedance		
							F08	Hospitalization or Prolonged Hospitalization	A071302	Inappropriate/Inadequate Shock/Stimulation		
									A0722	Impedance Problem		
									A072201	High Impedance		
							F2303	Medication Required	A071302	Inappropriate/Inadequate Shock/Stimulation		
									A0722	Impedance Problem		
									A072201	High Impedance		
7620	2022	All other reportable incidents	47267	Single-chamber implantable pacemaker, rate-responsive	CT141	Electrodes/Leads and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem	30/09/2022	30/09/2022
									A070504	Premature Discharge of Battery	30/09/2022	30/09/2022
							F08	Hospitalization or Prolonged Hospitalization	A0705	Battery Problem	30/09/2022	30/09/2022
									A070504	Premature Discharge of Battery	30/09/2022	30/09/2022
							F1905	Device Revision or Replacement	A0705	Battery Problem	30/09/2022	30/09/2022
									A070504	Premature Discharge of Battery	30/09/2022	30/09/2022
							F26	No Health Consequences or Impact	A0705	Battery Problem	30/09/2022	30/09/2022
									A070504	Premature Discharge of Battery	30/09/2022	30/09/2022
					CT145	Cardiovascular devices	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem	30/09/2022	30/09/2022
									A070504	Premature Discharge of Battery	30/09/2022	30/09/2022
							F08	Hospitalization or Prolonged Hospitalization	A0705	Battery Problem	30/09/2022	30/09/2022
									A070504	Premature Discharge of Battery	30/09/2022	30/09/2022
							F1905	Device Revision or Replacement	A0705	Battery Problem	30/09/2022	30/09/2022
									A070504	Premature Discharge of Battery	30/09/2022	30/09/2022
							F26	No Health Consequences or Impact	A0705	Battery Problem	30/09/2022	30/09/2022
									A070504	Premature Discharge of Battery	30/09/2022	30/09/2022
					CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem	30/09/2022	30/09/2022
									A070504	Premature Discharge of Battery	30/09/2022	30/09/2022
							F08	Hospitalization or Prolonged Hospitalization	A0705	Battery Problem	30/09/2022	30/09/2022
									A070504	Premature Discharge of Battery	30/09/2022	30/09/2022
							F1905	Device Revision or Replacement	A0705	Battery Problem	30/09/2022	30/09/2022
									A070504	Premature Discharge of Battery	30/09/2022	30/09/2022
							F26	No Health Consequences or Impact	A0705	Battery Problem	30/09/2022	30/09/2022
									A070504	Premature Discharge of Battery	30/09/2022	30/09/2022
					CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem	30/09/2022	30/09/2022
									A070504	Premature Discharge of Battery	30/09/2022	30/09/2022
							F08	Hospitalization or Prolonged Hospitalization	A0705	Battery Problem	30/09/2022	30/09/2022
									A070504	Premature Discharge of Battery	30/09/2022	30/09/2022
							F1905	Device Revision or Replacement	A0705	Battery Problem	30/09/2022	30/09/2022
									A070504	Premature Discharge of Battery	30/09/2022	30/09/2022
							F26	No Health Consequences or Impact	A0705	Battery Problem	30/09/2022	30/09/2022
									A070504	Premature Discharge of Battery	30/09/2022	30/09/2022
					CT233	Surgical	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem	30/09/2022	30/09/2022
									A070504	Premature Discharge of Battery	30/09/2022	30/09/2022
							F08	Hospitalization or Prolonged Hospitalization	A0705	Battery Problem	30/09/2022	30/09/2022
									A070504	Premature Discharge of Battery	30/09/2022	30/09/2022
							F1905	Device Revision or Replacement	A0705	Battery Problem	30/09/2022	30/09/2022
									A070504	Premature Discharge of Battery	30/09/2022	30/09/2022
							F26	No Health Consequences or Impact	A0705	Battery Problem	30/09/2022	30/09/2022
									A070504	Premature Discharge of Battery	30/09/2022	30/09/2022
					CT2406	Implantable	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem	30/09/2022	30/09/2022
									A070504	Premature Discharge of Battery	30/09/2022	30/09/2022
							F08	Hospitalization or Prolonged Hospitalization	A0705	Battery Problem	30/09/2022	30/09/2022
									A070504	Premature Discharge of Battery	30/09/2022	30/09/2022
							F1905	Device Revision or Replacement	A0705	Battery Problem	30/09/2022	30/09/2022
									A070504	Premature Discharge of Battery	30/09/2022	30/09/2022
							F26	No Health Consequences or Impact	A0705	Battery Problem	30/09/2022	30/09/2022
									A070504	Premature Discharge of Battery	30/09/2022	30/09/2022
					CT2407	Active implantable	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem	30/09/2022	30/09/2022
									A070504	Premature Discharge of Battery	30/09/2022	30/09/2022
							F08	Hospitalization or Prolonged Hospitalization	A0705	Battery Problem	30/09/2022	30/09/2022
									A070504	Premature Discharge of Battery	30/09/2022	30/09/2022
							F1905	Device Revision or Replacement	A0705	Battery Problem	30/09/2022	30/09/2022
									A070504	Premature Discharge of Battery	30/09/2022	30/09/2022
							F26	No Health Consequences or Impact	A0705	Battery Problem	30/09/2022	30/09/2022
									A070504	Premature Discharge of Battery	30/09/2022	30/09/2022
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem	30/09/2022	30/09/2022
									A070504	Premature Discharge of Battery	30/09/2022	30/09/2022

							F08	Hospitalization or Prolonged Hospitalization	A0705	Battery Problem	30/09/2022	30/09/2022
									A070504	Premature Discharge of Battery	30/09/2022	30/09/2022
							F1905	Device Revision or Replacement	A0705	Battery Problem	30/09/2022	30/09/2022
									A070504	Premature Discharge of Battery	30/09/2022	30/09/2022
							F26	No Health Consequences or Impact	A0705	Battery Problem	30/09/2022	30/09/2022
									A070504	Premature Discharge of Battery	30/09/2022	30/09/2022
					CT2727	Active implant devices/ancillaries	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem	30/09/2022	30/09/2022
									A070504	Premature Discharge of Battery	30/09/2022	30/09/2022
							F08	Hospitalization or Prolonged Hospitalization	A0705	Battery Problem	30/09/2022	30/09/2022
									A070504	Premature Discharge of Battery	30/09/2022	30/09/2022
							F1905	Device Revision or Replacement	A0705	Battery Problem	30/09/2022	30/09/2022
									A070504	Premature Discharge of Battery	30/09/2022	30/09/2022
							F26	No Health Consequences or Impact	A0705	Battery Problem	30/09/2022	30/09/2022
									A070504	Premature Discharge of Battery	30/09/2022	30/09/2022
					CT2750	Implantable cardiac electrical stimulators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem	30/09/2022	30/09/2022
									A070504	Premature Discharge of Battery	30/09/2022	30/09/2022
							F08	Hospitalization or Prolonged Hospitalization	A0705	Battery Problem	30/09/2022	30/09/2022
									A070504	Premature Discharge of Battery	30/09/2022	30/09/2022
							F1905	Device Revision or Replacement	A0705	Battery Problem	30/09/2022	30/09/2022
									A070504	Premature Discharge of Battery	30/09/2022	30/09/2022
							F26	No Health Consequences or Impact	A0705	Battery Problem	30/09/2022	30/09/2022
									A070504	Premature Discharge of Battery	30/09/2022	30/09/2022
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem	30/09/2022	30/09/2022
									A070504	Premature Discharge of Battery	30/09/2022	30/09/2022
							F08	Hospitalization or Prolonged Hospitalization	A0705	Battery Problem	30/09/2022	30/09/2022
									A070504	Premature Discharge of Battery	30/09/2022	30/09/2022
							F1905	Device Revision or Replacement	A0705	Battery Problem	30/09/2022	30/09/2022
									A070504	Premature Discharge of Battery	30/09/2022	30/09/2022
							F26	No Health Consequences or Impact	A0705	Battery Problem	30/09/2022	30/09/2022
									A070504	Premature Discharge of Battery	30/09/2022	30/09/2022
					CT2985	Active	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem	30/09/2022	30/09/2022
									A070504	Premature Discharge of Battery	30/09/2022	30/09/2022
							F08	Hospitalization or Prolonged Hospitalization	A0705	Battery Problem	30/09/2022	30/09/2022
									A070504	Premature Discharge of Battery	30/09/2022	30/09/2022
							F1905	Device Revision or Replacement	A0705	Battery Problem	30/09/2022	30/09/2022
									A070504	Premature Discharge of Battery	30/09/2022	30/09/2022
							F26	No Health Consequences or Impact	A0705	Battery Problem	30/09/2022	30/09/2022
									A070504	Premature Discharge of Battery	30/09/2022	30/09/2022
					CT3001	Implantable metal	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem	30/09/2022	30/09/2022
									A070504	Premature Discharge of Battery	30/09/2022	30/09/2022
							F08	Hospitalization or Prolonged Hospitalization	A0705	Battery Problem	30/09/2022	30/09/2022
									A070504	Premature Discharge of Battery	30/09/2022	30/09/2022
							F1905	Device Revision or Replacement	A0705	Battery Problem	30/09/2022	30/09/2022
									A070504	Premature Discharge of Battery	30/09/2022	30/09/2022
							F26	No Health Consequences or Impact	A0705	Battery Problem	30/09/2022	30/09/2022
									A070504	Premature Discharge of Battery	30/09/2022	30/09/2022
					CT3002	Implantable synthetic polymer	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem	30/09/2022	30/09/2022
									A070504	Premature Discharge of Battery	30/09/2022	30/09/2022
							F08	Hospitalization or Prolonged Hospitalization	A0705	Battery Problem	30/09/2022	30/09/2022
									A070504	Premature Discharge of Battery	30/09/2022	30/09/2022
							F1905	Device Revision or Replacement	A0705	Battery Problem	30/09/2022	30/09/2022
									A070504	Premature Discharge of Battery	30/09/2022	30/09/2022
							F26	No Health Consequences or Impact	A0705	Battery Problem	30/09/2022	30/09/2022
									A070504	Premature Discharge of Battery	30/09/2022	30/09/2022
					CT3006	Implantable inorganic material	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem	30/09/2022	30/09/2022
									A070504	Premature Discharge of Battery	30/09/2022	30/09/2022
							F08	Hospitalization or Prolonged Hospitalization	A0705	Battery Problem	30/09/2022	30/09/2022
									A070504	Premature Discharge of Battery	30/09/2022	30/09/2022
							F1905	Device Revision or Replacement	A0705	Battery Problem	30/09/2022	30/09/2022
									A070504	Premature Discharge of Battery	30/09/2022	30/09/2022
							F26	No Health Consequences or Impact	A0705	Battery Problem	30/09/2022	30/09/2022
									A070504	Premature Discharge of Battery	30/09/2022	30/09/2022
					CT314	Electrically powered	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem	30/09/2022	30/09/2022
									A070504	Premature Discharge of Battery	30/09/2022	30/09/2022
							F08	Hospitalization or Prolonged Hospitalization	A0705	Battery Problem	30/09/2022	30/09/2022
									A070504	Premature Discharge of Battery	30/09/2022	30/09/2022
							F1905	Device Revision or Replacement	A0705	Battery Problem	30/09/2022	30/09/2022
									A070504	Premature Discharge of Battery	30/09/2022	30/09/2022
							F26	No Health Consequences or Impact	A0705	Battery Problem	30/09/2022	30/09/2022
									A070504	Premature Discharge of Battery	30/09/2022	30/09/2022
					CT321	Long-term surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem	30/09/2022	30/09/2022
									A070504	Premature Discharge of Battery	30/09/2022	30/09/2022
							F08	Hospitalization or Prolonged Hospitalization	A0705	Battery Problem	30/09/2022	30/09/2022
									A070504	Premature Discharge of Battery	30/09/2022	30/09/2022
							F1905	Device Revision or Replacement	A0705	Battery Problem	30/09/2022	30/09/2022
									A070504	Premature Discharge of Battery	30/09/2022	30/09/2022
							F26	No Health Consequences or Impact	A0705	Battery Problem	30/09/2022	30/09/2022
									A070504	Premature Discharge of Battery	30/09/2022	30/09/2022
					CT449	Generators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem	30/09/2022	30/09/2022
									A070504	Premature Discharge of Battery	30/09/2022	30/09/2022
							F08	Hospitalization or Prolonged Hospitalization	A0705	Battery Problem	30/09/2022	30/09/2022
									A070504	Premature Discharge of Battery	30/09/2022	30/09/2022
							F1905	Device Revision or Replacement	A0705	Battery Problem	30/09/2022	30/09/2022
									A070504	Premature Discharge of Battery	30/09/2022	30/09/2022
							F26	No Health Consequences or Impact	A0705	Battery Problem	30/09/2022	30/09/2022
									A070504	Premature Discharge of Battery	30/09/2022	30/09/2022
					CT981	Single use	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem	30/09/2022	30/09/2022
									A070504	Premature Discharge of Battery	30/09/2022	30/09/2022

							F08	Hospitalization or Prolonged Hospitalization	A0705	Battery Problem	30/09/2022	30/09/2022
									A070504	Premature Discharge of Battery	30/09/2022	30/09/2022
							F1905	Device Revision or Replacement	A0705	Battery Problem	30/09/2022	30/09/2022
									A070504	Premature Discharge of Battery	30/09/2022	30/09/2022
							F26	No Health Consequences or Impact	A0705	Battery Problem	30/09/2022	30/09/2022
									A070504	Premature Discharge of Battery	30/09/2022	30/09/2022
					CT983	Surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem	30/09/2022	30/09/2022
									A070504	Premature Discharge of Battery	30/09/2022	30/09/2022
							F08	Hospitalization or Prolonged Hospitalization	A0705	Battery Problem	30/09/2022	30/09/2022
									A070504	Premature Discharge of Battery	30/09/2022	30/09/2022
							F1905	Device Revision or Replacement	A0705	Battery Problem	30/09/2022	30/09/2022
									A070504	Premature Discharge of Battery	30/09/2022	30/09/2022
							F26	No Health Consequences or Impact	A0705	Battery Problem	30/09/2022	30/09/2022
									A070504	Premature Discharge of Battery	30/09/2022	30/09/2022
7621	2022	All other reportable incidents	37265	Dual-chamber implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A072201	High Impedance		
							F26	No Health Consequences or Impact	A072201	High Impedance		
7622	2022	All other reportable incidents	35852	Single-chamber implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem	11/01/2022	11/01/2022
									A070504	Premature Discharge of Battery	11/01/2022	11/01/2022
							F08	Hospitalization or Prolonged Hospitalization	A0705	Battery Problem	11/01/2022	11/01/2022
									A070504	Premature Discharge of Battery	11/01/2022	11/01/2022
							F1905	Device Revision or Replacement	A0705	Battery Problem	11/01/2022	11/01/2022
									A070504	Premature Discharge of Battery	11/01/2022	11/01/2022
7623	2022	All other reportable incidents	37265	Dual-chamber implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
7624	2022	All other reportable incidents	35852	Single-chamber implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E2104	Electric Shock	A0709	Device Sensing Problem		
									A070909	Oversensing		
									A071302	Inappropriate/Inadequate Shock/Stimulation		
									A1301	Failure to Read Input Signal		
							E2403	No Clinical Signs, Symptoms or Conditions	A0709	Device Sensing Problem		
									A070909	Oversensing		
									A071302	Inappropriate/Inadequate Shock/Stimulation		
									A1301	Failure to Read Input Signal		
							F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A0709	Device Sensing Problem		
									A070909	Oversensing		
									A071302	Inappropriate/Inadequate Shock/Stimulation		
									A1301	Failure to Read Input Signal		
7625	2022	All other reportable incidents	36241	Pacemaker/defibrillator lead	CT1684	Cardiac pacemakers and associated devices	E2104	Electric Shock	A0709	Device Sensing Problem		
									A070909	Oversensing		
									A071302	Inappropriate/Inadequate Shock/Stimulation		
									A1301	Failure to Read Input Signal		
							E2403	No Clinical Signs, Symptoms or Conditions	A0709	Device Sensing Problem		
									A070909	Oversensing		
									A071302	Inappropriate/Inadequate Shock/Stimulation		
									A1301	Failure to Read Input Signal		
							F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A0709	Device Sensing Problem		
									A070909	Oversensing		
									A071302	Inappropriate/Inadequate Shock/Stimulation		
									A1301	Failure to Read Input Signal		
7626	2022	All other reportable incidents	35852	Single-chamber implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A071201	Failure to Convert Rhythm		
							F26	No Health Consequences or Impact	A071201	Failure to Convert Rhythm		
7627	2022	All other reportable incidents	47270	Cardiac resynchronization therapy implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E0601	Arrhythmia	A070909	Oversensing		
									A071204	No Pacing		
							E2403	No Clinical Signs, Symptoms or Conditions	A070909	Oversensing		
									A071204	No Pacing		
							F23	Unexpected Medical Intervention	A070909	Oversensing		
									A071204	No Pacing		
7628	2022	All other reportable incidents	17882	Physiologic-monitoring defibrillation system	CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A1603	Failure of Device to Self-Test		
							F27	Problem Identified During Non-Clinical Procedure	A1603	Failure of Device to Self-Test		
7629	2022	All other reportable incidents	47267	Single-chamber implantable pacemaker, rate-responsive	CT141	Electrodes/Leads and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem	06/09/2022	06/09/2022
									A070504	Premature Discharge of Battery	06/09/2022	06/09/2022
							F08	Hospitalization or Prolonged Hospitalization	A0705	Battery Problem	06/09/2022	06/09/2022
									A070504	Premature Discharge of Battery	06/09/2022	06/09/2022
							F1905	Device Revision or Replacement	A0705	Battery Problem	06/09/2022	06/09/2022
									A070504	Premature Discharge of Battery	06/09/2022	06/09/2022
							F26	No Health Consequences or Impact	A0705	Battery Problem	06/09/2022	06/09/2022
									A070504	Premature Discharge of Battery	06/09/2022	06/09/2022
					CT145	Cardiovascular devices	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem	06/09/2022	06/09/2022
									A070504	Premature Discharge of Battery	06/09/2022	06/09/2022
							F08	Hospitalization or Prolonged Hospitalization	A0705	Battery Problem	06/09/2022	06/09/2022
									A070504	Premature Discharge of Battery	06/09/2022	06/09/2022
							F1905	Device Revision or Replacement	A0705	Battery Problem	06/09/2022	06/09/2022
									A070504	Premature Discharge of Battery	06/09/2022	06/09/2022
							F26	No Health Consequences or Impact	A0705	Battery Problem	06/09/2022	06/09/2022
									A070504	Premature Discharge of Battery	06/09/2022	06/09/2022
					CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem	06/09/2022	06/09/2022
									A070504	Premature Discharge of Battery	06/09/2022	06/09/2022
							F08	Hospitalization or Prolonged Hospitalization	A0705	Battery Problem	06/09/2022	06/09/2022
									A070504	Premature Discharge of Battery	06/09/2022	06/09/2022
							F1905	Device Revision or Replacement	A0705	Battery Problem	06/09/2022	06/09/2022
									A070504	Premature Discharge of Battery	06/09/2022	06/09/2022
							F26	No Health Consequences or Impact	A0705	Battery Problem	06/09/2022	06/09/2022
									A070504	Premature Discharge of Battery	06/09/2022	06/09/2022
					CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem	06/09/2022	06/09/2022
									A070504	Premature Discharge of Battery	06/09/2022	06/09/2022
							F08	Hospitalization or Prolonged Hospitalization	A0705	Battery Problem	06/09/2022	06/09/2022
									A070504	Premature Discharge of Battery	06/09/2022	06/09/2022
							F1905	Device Revision or Replacement	A0705	Battery Problem	06/09/2022	06/09/2022
									A070504	Premature Discharge of Battery	06/09/2022	06/09/2022

							F26	No Health Consequences or Impact	A0705	Battery Problem	06/09/2022	06/09/2022
									A070504	Premature Discharge of Battery	06/09/2022	06/09/2022
					CT233	Surgical	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem	06/09/2022	06/09/2022
									A070504	Premature Discharge of Battery	06/09/2022	06/09/2022
							F08	Hospitalization or Prolonged Hospitalization	A0705	Battery Problem	06/09/2022	06/09/2022
									A070504	Premature Discharge of Battery	06/09/2022	06/09/2022
							F1905	Device Revision or Replacement	A0705	Battery Problem	06/09/2022	06/09/2022
									A070504	Premature Discharge of Battery	06/09/2022	06/09/2022
							F26	No Health Consequences or Impact	A0705	Battery Problem	06/09/2022	06/09/2022
									A070504	Premature Discharge of Battery	06/09/2022	06/09/2022
					CT2406	Implantable	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem	06/09/2022	06/09/2022
									A070504	Premature Discharge of Battery	06/09/2022	06/09/2022
							F08	Hospitalization or Prolonged Hospitalization	A0705	Battery Problem	06/09/2022	06/09/2022
									A070504	Premature Discharge of Battery	06/09/2022	06/09/2022
							F1905	Device Revision or Replacement	A0705	Battery Problem	06/09/2022	06/09/2022
									A070504	Premature Discharge of Battery	06/09/2022	06/09/2022
							F26	No Health Consequences or Impact	A0705	Battery Problem	06/09/2022	06/09/2022
									A070504	Premature Discharge of Battery	06/09/2022	06/09/2022
					CT2407	Active implantable	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem	06/09/2022	06/09/2022
									A070504	Premature Discharge of Battery	06/09/2022	06/09/2022
							F08	Hospitalization or Prolonged Hospitalization	A0705	Battery Problem	06/09/2022	06/09/2022
									A070504	Premature Discharge of Battery	06/09/2022	06/09/2022
							F1905	Device Revision or Replacement	A0705	Battery Problem	06/09/2022	06/09/2022
									A070504	Premature Discharge of Battery	06/09/2022	06/09/2022
							F26	No Health Consequences or Impact	A0705	Battery Problem	06/09/2022	06/09/2022
									A070504	Premature Discharge of Battery	06/09/2022	06/09/2022
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem	06/09/2022	06/09/2022
									A070504	Premature Discharge of Battery	06/09/2022	06/09/2022
							F08	Hospitalization or Prolonged Hospitalization	A0705	Battery Problem	06/09/2022	06/09/2022
									A070504	Premature Discharge of Battery	06/09/2022	06/09/2022
							F1905	Device Revision or Replacement	A0705	Battery Problem	06/09/2022	06/09/2022
									A070504	Premature Discharge of Battery	06/09/2022	06/09/2022
							F26	No Health Consequences or Impact	A0705	Battery Problem	06/09/2022	06/09/2022
									A070504	Premature Discharge of Battery	06/09/2022	06/09/2022
					CT2727	Active implant devices/ancillaries	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem	06/09/2022	06/09/2022
									A070504	Premature Discharge of Battery	06/09/2022	06/09/2022
							F08	Hospitalization or Prolonged Hospitalization	A0705	Battery Problem	06/09/2022	06/09/2022
									A070504	Premature Discharge of Battery	06/09/2022	06/09/2022
							F1905	Device Revision or Replacement	A0705	Battery Problem	06/09/2022	06/09/2022
									A070504	Premature Discharge of Battery	06/09/2022	06/09/2022
							F26	No Health Consequences or Impact	A0705	Battery Problem	06/09/2022	06/09/2022
									A070504	Premature Discharge of Battery	06/09/2022	06/09/2022
					CT2750	Implantable cardiac electrical stimulators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem	06/09/2022	06/09/2022
									A070504	Premature Discharge of Battery	06/09/2022	06/09/2022
							F08	Hospitalization or Prolonged Hospitalization	A0705	Battery Problem	06/09/2022	06/09/2022
									A070504	Premature Discharge of Battery	06/09/2022	06/09/2022
							F1905	Device Revision or Replacement	A0705	Battery Problem	06/09/2022	06/09/2022
									A070504	Premature Discharge of Battery	06/09/2022	06/09/2022
							F26	No Health Consequences or Impact	A0705	Battery Problem	06/09/2022	06/09/2022
									A070504	Premature Discharge of Battery	06/09/2022	06/09/2022
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem	06/09/2022	06/09/2022
									A070504	Premature Discharge of Battery	06/09/2022	06/09/2022
							F08	Hospitalization or Prolonged Hospitalization	A0705	Battery Problem	06/09/2022	06/09/2022
									A070504	Premature Discharge of Battery	06/09/2022	06/09/2022
							F1905	Device Revision or Replacement	A0705	Battery Problem	06/09/2022	06/09/2022
									A070504	Premature Discharge of Battery	06/09/2022	06/09/2022
							F26	No Health Consequences or Impact	A0705	Battery Problem	06/09/2022	06/09/2022
									A070504	Premature Discharge of Battery	06/09/2022	06/09/2022
					CT2985	Active	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem	06/09/2022	06/09/2022
									A070504	Premature Discharge of Battery	06/09/2022	06/09/2022
							F08	Hospitalization or Prolonged Hospitalization	A0705	Battery Problem	06/09/2022	06/09/2022
									A070504	Premature Discharge of Battery	06/09/2022	06/09/2022
							F1905	Device Revision or Replacement	A0705	Battery Problem	06/09/2022	06/09/2022
									A070504	Premature Discharge of Battery	06/09/2022	06/09/2022
							F26	No Health Consequences or Impact	A0705	Battery Problem	06/09/2022	06/09/2022
									A070504	Premature Discharge of Battery	06/09/2022	06/09/2022
					CT3001	Implantable metal	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem	06/09/2022	06/09/2022
									A070504	Premature Discharge of Battery	06/09/2022	06/09/2022
							F08	Hospitalization or Prolonged Hospitalization	A0705	Battery Problem	06/09/2022	06/09/2022
									A070504	Premature Discharge of Battery	06/09/2022	06/09/2022
							F1905	Device Revision or Replacement	A0705	Battery Problem	06/09/2022	06/09/2022
									A070504	Premature Discharge of Battery	06/09/2022	06/09/2022
							F26	No Health Consequences or Impact	A0705	Battery Problem	06/09/2022	06/09/2022
									A070504	Premature Discharge of Battery	06/09/2022	06/09/2022
					CT3002	Implantable synthetic polymer	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem	06/09/2022	06/09/2022
									A070504	Premature Discharge of Battery	06/09/2022	06/09/2022
							F08	Hospitalization or Prolonged Hospitalization	A0705	Battery Problem	06/09/2022	06/09/2022
									A070504	Premature Discharge of Battery	06/09/2022	06/09/2022
							F1905	Device Revision or Replacement	A0705	Battery Problem	06/09/2022	06/09/2022
									A070504	Premature Discharge of Battery	06/09/2022	06/09/2022
							F26	No Health Consequences or Impact	A0705	Battery Problem	06/09/2022	06/09/2022
									A070504	Premature Discharge of Battery	06/09/2022	06/09/2022
					CT3006	Implantable inorganic material	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem	06/09/2022	06/09/2022
									A070504	Premature Discharge of Battery	06/09/2022	06/09/2022
							F08	Hospitalization or Prolonged Hospitalization	A0705	Battery Problem	06/09/2022	06/09/2022
									A070504	Premature Discharge of Battery	06/09/2022	06/09/2022
							F1905	Device Revision or Replacement	A0705	Battery Problem	06/09/2022	06/09/2022
									A070504	Premature Discharge of Battery	06/09/2022	06/09/2022

							F26	No Health Consequences or Impact	A0705	Battery Problem	06/09/2022	06/09/2022
									A070504	Premature Discharge of Battery	06/09/2022	06/09/2022
					CT314	Electrically powered	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem	06/09/2022	06/09/2022
									A070504	Premature Discharge of Battery	06/09/2022	06/09/2022
							F08	Hospitalization or Prolonged Hospitalization	A0705	Battery Problem	06/09/2022	06/09/2022
									A070504	Premature Discharge of Battery	06/09/2022	06/09/2022
							F1905	Device Revision or Replacement	A0705	Battery Problem	06/09/2022	06/09/2022
									A070504	Premature Discharge of Battery	06/09/2022	06/09/2022
							F26	No Health Consequences or Impact	A0705	Battery Problem	06/09/2022	06/09/2022
									A070504	Premature Discharge of Battery	06/09/2022	06/09/2022
					CT321	Long-term surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem	06/09/2022	06/09/2022
									A070504	Premature Discharge of Battery	06/09/2022	06/09/2022
							F08	Hospitalization or Prolonged Hospitalization	A0705	Battery Problem	06/09/2022	06/09/2022
									A070504	Premature Discharge of Battery	06/09/2022	06/09/2022
							F1905	Device Revision or Replacement	A0705	Battery Problem	06/09/2022	06/09/2022
									A070504	Premature Discharge of Battery	06/09/2022	06/09/2022
							F26	No Health Consequences or Impact	A0705	Battery Problem	06/09/2022	06/09/2022
									A070504	Premature Discharge of Battery	06/09/2022	06/09/2022
					CT449	Generators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem	06/09/2022	06/09/2022
									A070504	Premature Discharge of Battery	06/09/2022	06/09/2022
							F08	Hospitalization or Prolonged Hospitalization	A0705	Battery Problem	06/09/2022	06/09/2022
									A070504	Premature Discharge of Battery	06/09/2022	06/09/2022
							F1905	Device Revision or Replacement	A0705	Battery Problem	06/09/2022	06/09/2022
									A070504	Premature Discharge of Battery	06/09/2022	06/09/2022
							F26	No Health Consequences or Impact	A0705	Battery Problem	06/09/2022	06/09/2022
									A070504	Premature Discharge of Battery	06/09/2022	06/09/2022
					CT981	Single use	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem	06/09/2022	06/09/2022
									A070504	Premature Discharge of Battery	06/09/2022	06/09/2022
							F08	Hospitalization or Prolonged Hospitalization	A0705	Battery Problem	06/09/2022	06/09/2022
									A070504	Premature Discharge of Battery	06/09/2022	06/09/2022
							F1905	Device Revision or Replacement	A0705	Battery Problem	06/09/2022	06/09/2022
									A070504	Premature Discharge of Battery	06/09/2022	06/09/2022
							F26	No Health Consequences or Impact	A0705	Battery Problem	06/09/2022	06/09/2022
									A070504	Premature Discharge of Battery	06/09/2022	06/09/2022
					CT983	Surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem	06/09/2022	06/09/2022
									A070504	Premature Discharge of Battery	06/09/2022	06/09/2022
							F08	Hospitalization or Prolonged Hospitalization	A0705	Battery Problem	06/09/2022	06/09/2022
									A070504	Premature Discharge of Battery	06/09/2022	06/09/2022
							F1905	Device Revision or Replacement	A0705	Battery Problem	06/09/2022	06/09/2022
									A070504	Premature Discharge of Battery	06/09/2022	06/09/2022
							F26	No Health Consequences or Impact	A0705	Battery Problem	06/09/2022	06/09/2022
									A070504	Premature Discharge of Battery	06/09/2022	06/09/2022
7630	2022	All other reportable incidents	47270	Cardiac resynchronization therapy implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
7631	2022	All other reportable incidents	36241	Pacemaker/defibrillator lead	CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070101	Failure to Capture		
									A072201	High Impedance		
							F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A070101	Failure to Capture		
									A072201	High Impedance		
7632	2022	All other reportable incidents	47267	Single-chamber implantable pacemaker, rate-responsive	CT141	Electrodes/Leads and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem	01/03/2022	01/03/2022
									A070504	Premature Discharge of Battery	01/03/2022	01/03/2022
							F08	Hospitalization or Prolonged Hospitalization	A0705	Battery Problem	01/03/2022	01/03/2022
									A070504	Premature Discharge of Battery	01/03/2022	01/03/2022
							F1905	Device Revision or Replacement	A0705	Battery Problem	01/03/2022	01/03/2022
									A070504	Premature Discharge of Battery	01/03/2022	01/03/2022
					CT145	Cardiovascular devices	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem	01/03/2022	01/03/2022
									A070504	Premature Discharge of Battery	01/03/2022	01/03/2022
							F08	Hospitalization or Prolonged Hospitalization	A0705	Battery Problem	01/03/2022	01/03/2022
									A070504	Premature Discharge of Battery	01/03/2022	01/03/2022
							F1905	Device Revision or Replacement	A0705	Battery Problem	01/03/2022	01/03/2022
									A070504	Premature Discharge of Battery	01/03/2022	01/03/2022
					CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem	01/03/2022	01/03/2022
									A070504	Premature Discharge of Battery	01/03/2022	01/03/2022
							F08	Hospitalization or Prolonged Hospitalization	A0705	Battery Problem	01/03/2022	01/03/2022
									A070504	Premature Discharge of Battery	01/03/2022	01/03/2022
							F1905	Device Revision or Replacement	A0705	Battery Problem	01/03/2022	01/03/2022
									A070504	Premature Discharge of Battery	01/03/2022	01/03/2022
					CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem	01/03/2022	01/03/2022
									A070504	Premature Discharge of Battery	01/03/2022	01/03/2022
							F08	Hospitalization or Prolonged Hospitalization	A0705	Battery Problem	01/03/2022	01/03/2022
									A070504	Premature Discharge of Battery	01/03/2022	01/03/2022
							F1905	Device Revision or Replacement	A0705	Battery Problem	01/03/2022	01/03/2022
									A070504	Premature Discharge of Battery	01/03/2022	01/03/2022
					CT233	Surgical	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem	01/03/2022	01/03/2022
									A070504	Premature Discharge of Battery	01/03/2022	01/03/2022
							F08	Hospitalization or Prolonged Hospitalization	A0705	Battery Problem	01/03/2022	01/03/2022
									A070504	Premature Discharge of Battery	01/03/2022	01/03/2022
							F1905	Device Revision or Replacement	A0705	Battery Problem	01/03/2022	01/03/2022
									A070504	Premature Discharge of Battery	01/03/2022	01/03/2022
					CT2406	Implantable	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem	01/03/2022	01/03/2022
									A070504	Premature Discharge of Battery	01/03/2022	01/03/2022
							F08	Hospitalization or Prolonged Hospitalization	A0705	Battery Problem	01/03/2022	01/03/2022
									A070504	Premature Discharge of Battery	01/03/2022	01/03/2022
							F1905	Device Revision or Replacement	A0705	Battery Problem	01/03/2022	01/03/2022
									A070504	Premature Discharge of Battery	01/03/2022	01/03/2022
					CT2407	Active implantable	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem	01/03/2022	01/03/2022
									A070504	Premature Discharge of Battery	01/03/2022	01/03/2022
							F08	Hospitalization or Prolonged Hospitalization	A0705	Battery Problem	01/03/2022	01/03/2022
									A070504	Premature Discharge of Battery	01/03/2022	01/03/2022

							F1905	Device Revision or Replacement	A0705	Battery Problem	01/03/2022	01/03/2022
									A070504	Premature Discharge of Battery	01/03/2022	01/03/2022
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem	01/03/2022	01/03/2022
									A070504	Premature Discharge of Battery	01/03/2022	01/03/2022
							F08	Hospitalization or Prolonged Hospitalization	A0705	Battery Problem	01/03/2022	01/03/2022
									A070504	Premature Discharge of Battery	01/03/2022	01/03/2022
							F1905	Device Revision or Replacement	A0705	Battery Problem	01/03/2022	01/03/2022
									A070504	Premature Discharge of Battery	01/03/2022	01/03/2022
					CT2727	Active implant devices/ancillaries	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem	01/03/2022	01/03/2022
									A070504	Premature Discharge of Battery	01/03/2022	01/03/2022
							F08	Hospitalization or Prolonged Hospitalization	A0705	Battery Problem	01/03/2022	01/03/2022
									A070504	Premature Discharge of Battery	01/03/2022	01/03/2022
							F1905	Device Revision or Replacement	A0705	Battery Problem	01/03/2022	01/03/2022
									A070504	Premature Discharge of Battery	01/03/2022	01/03/2022
					CT2750	Implantable cardiac electrical stimulators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem	01/03/2022	01/03/2022
									A070504	Premature Discharge of Battery	01/03/2022	01/03/2022
							F08	Hospitalization or Prolonged Hospitalization	A0705	Battery Problem	01/03/2022	01/03/2022
									A070504	Premature Discharge of Battery	01/03/2022	01/03/2022
							F1905	Device Revision or Replacement	A0705	Battery Problem	01/03/2022	01/03/2022
									A070504	Premature Discharge of Battery	01/03/2022	01/03/2022
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem	01/03/2022	01/03/2022
									A070504	Premature Discharge of Battery	01/03/2022	01/03/2022
							F08	Hospitalization or Prolonged Hospitalization	A0705	Battery Problem	01/03/2022	01/03/2022
									A070504	Premature Discharge of Battery	01/03/2022	01/03/2022
							F1905	Device Revision or Replacement	A0705	Battery Problem	01/03/2022	01/03/2022
									A070504	Premature Discharge of Battery	01/03/2022	01/03/2022
					CT2985	Active	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem	01/03/2022	01/03/2022
									A070504	Premature Discharge of Battery	01/03/2022	01/03/2022
							F08	Hospitalization or Prolonged Hospitalization	A0705	Battery Problem	01/03/2022	01/03/2022
									A070504	Premature Discharge of Battery	01/03/2022	01/03/2022
							F1905	Device Revision or Replacement	A0705	Battery Problem	01/03/2022	01/03/2022
									A070504	Premature Discharge of Battery	01/03/2022	01/03/2022
					CT3001	Implantable metal	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem	01/03/2022	01/03/2022
									A070504	Premature Discharge of Battery	01/03/2022	01/03/2022
							F08	Hospitalization or Prolonged Hospitalization	A0705	Battery Problem	01/03/2022	01/03/2022
									A070504	Premature Discharge of Battery	01/03/2022	01/03/2022
							F1905	Device Revision or Replacement	A0705	Battery Problem	01/03/2022	01/03/2022
									A070504	Premature Discharge of Battery	01/03/2022	01/03/2022
					CT3002	Implantable synthetic polymer	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem	01/03/2022	01/03/2022
									A070504	Premature Discharge of Battery	01/03/2022	01/03/2022
							F08	Hospitalization or Prolonged Hospitalization	A0705	Battery Problem	01/03/2022	01/03/2022
									A070504	Premature Discharge of Battery	01/03/2022	01/03/2022
							F1905	Device Revision or Replacement	A0705	Battery Problem	01/03/2022	01/03/2022
									A070504	Premature Discharge of Battery	01/03/2022	01/03/2022
					CT3006	Implantable inorganic material	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem	01/03/2022	01/03/2022
									A070504	Premature Discharge of Battery	01/03/2022	01/03/2022
							F08	Hospitalization or Prolonged Hospitalization	A0705	Battery Problem	01/03/2022	01/03/2022
									A070504	Premature Discharge of Battery	01/03/2022	01/03/2022
							F1905	Device Revision or Replacement	A0705	Battery Problem	01/03/2022	01/03/2022
									A070504	Premature Discharge of Battery	01/03/2022	01/03/2022
					CT314	Electrically powered	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem	01/03/2022	01/03/2022
									A070504	Premature Discharge of Battery	01/03/2022	01/03/2022
							F08	Hospitalization or Prolonged Hospitalization	A0705	Battery Problem	01/03/2022	01/03/2022
									A070504	Premature Discharge of Battery	01/03/2022	01/03/2022
							F1905	Device Revision or Replacement	A0705	Battery Problem	01/03/2022	01/03/2022
									A070504	Premature Discharge of Battery	01/03/2022	01/03/2022
					CT321	Long-term surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem	01/03/2022	01/03/2022
									A070504	Premature Discharge of Battery	01/03/2022	01/03/2022
							F08	Hospitalization or Prolonged Hospitalization	A0705	Battery Problem	01/03/2022	01/03/2022
									A070504	Premature Discharge of Battery	01/03/2022	01/03/2022
							F1905	Device Revision or Replacement	A0705	Battery Problem	01/03/2022	01/03/2022
									A070504	Premature Discharge of Battery	01/03/2022	01/03/2022
					CT449	Generators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem	01/03/2022	01/03/2022
									A070504	Premature Discharge of Battery	01/03/2022	01/03/2022
							F08	Hospitalization or Prolonged Hospitalization	A0705	Battery Problem	01/03/2022	01/03/2022
									A070504	Premature Discharge of Battery	01/03/2022	01/03/2022
							F1905	Device Revision or Replacement	A0705	Battery Problem	01/03/2022	01/03/2022
									A070504	Premature Discharge of Battery	01/03/2022	01/03/2022
					CT981	Single use	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem	01/03/2022	01/03/2022
									A070504	Premature Discharge of Battery	01/03/2022	01/03/2022
							F08	Hospitalization or Prolonged Hospitalization	A0705	Battery Problem	01/03/2022	01/03/2022
									A070504	Premature Discharge of Battery	01/03/2022	01/03/2022
							F1905	Device Revision or Replacement	A0705	Battery Problem	01/03/2022	01/03/2022
									A070504	Premature Discharge of Battery	01/03/2022	01/03/2022
					CT983	Surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem	01/03/2022	01/03/2022
									A070504	Premature Discharge of Battery	01/03/2022	01/03/2022
							F08	Hospitalization or Prolonged Hospitalization	A0705	Battery Problem	01/03/2022	01/03/2022
									A070504	Premature Discharge of Battery	01/03/2022	01/03/2022
							F1905	Device Revision or Replacement	A0705	Battery Problem	01/03/2022	01/03/2022
									A070504	Premature Discharge of Battery	01/03/2022	01/03/2022
7633	2022	All other reportable incidents	47267	Single-chamber implantable pacemaker, rate-responsive	CT141	Electrodes/Leads and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
					CT145	Cardiovascular devices	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		

					CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
					CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
					CT233	Surgical	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
					CT2406	Implantable	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
					CT2407	Active implantable	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
					CT2727	Active implant devices/ancillaries	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
					CT2750	Implantable cardiac electrical stimulators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
					CT2985	Active	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
					CT3001	Implantable metal	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
					CT3002	Implantable synthetic polymer	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
					CT3006	Implantable inorganic material	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
					CT314	Electrically powered	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
					CT321	Long-term surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
					CT449	Generators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
					CT981	Single use	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
					CT983	Surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
7634	2022	All other reportable incidents	35223	Endocardial/interventricular septal pacing lead	CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070103	Intermittent Capture		
									A072201	High Impedance		
							F26	No Health Consequences or Impact	A070103	Intermittent Capture		
									A072201	High Impedance		
7635	2022	All other reportable incidents	35852	Single-chamber implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem	15/09/2022	15/09/2022
									A070504	Premature Discharge of Battery	15/09/2022	15/09/2022
							F08	Hospitalization or Prolonged Hospitalization	A0705	Battery Problem	15/09/2022	15/09/2022
									A070504	Premature Discharge of Battery	15/09/2022	15/09/2022
							F1905	Device Revision or Replacement	A0705	Battery Problem	15/09/2022	15/09/2022
									A070504	Premature Discharge of Battery	15/09/2022	15/09/2022
							F24	Insufficient Health Impact Information	A0705	Battery Problem	15/09/2022	15/09/2022
									A070504	Premature Discharge of Battery	15/09/2022	15/09/2022
7636	2022	All other reportable incidents	37265	Dual-chamber implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	26/01/2022	26/01/2022
							F1901	Additional Surgery	A070504	Premature Discharge of Battery	26/01/2022	26/01/2022
7637	2022	All other reportable incidents	47265	Dual-chamber implantable pacemaker, rate-responsive	CT141	Electrodes/Leads and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070101	Failure to Capture		
									A072201	High Impedance		

							F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A070101	Failure to Capture		
									A072201	High Impedance		
					CT145	Cardiovascular devices	E2403	No Clinical Signs, Symptoms or Conditions	A070101	Failure to Capture		
									A072201	High Impedance		
							F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A070101	Failure to Capture		
									A072201	High Impedance		
					CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070101	Failure to Capture		
									A072201	High Impedance		
							F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A070101	Failure to Capture		
									A072201	High Impedance		
					CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A070101	Failure to Capture		
									A072201	High Impedance		
							F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A070101	Failure to Capture		
									A072201	High Impedance		
					CT233	Surgical	E2403	No Clinical Signs, Symptoms or Conditions	A070101	Failure to Capture		
									A072201	High Impedance		
							F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A070101	Failure to Capture		
									A072201	High Impedance		
					CT2406	Implantable	E2403	No Clinical Signs, Symptoms or Conditions	A070101	Failure to Capture		
									A072201	High Impedance		
							F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A070101	Failure to Capture		
									A072201	High Impedance		
					CT2407	Active implantable	E2403	No Clinical Signs, Symptoms or Conditions	A070101	Failure to Capture		
									A072201	High Impedance		
							F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A070101	Failure to Capture		
									A072201	High Impedance		
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A070101	Failure to Capture		
									A072201	High Impedance		
							F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A070101	Failure to Capture		
									A072201	High Impedance		
					CT2727	Active implant devices/ancillaries	E2403	No Clinical Signs, Symptoms or Conditions	A070101	Failure to Capture		
									A072201	High Impedance		
							F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A070101	Failure to Capture		
									A072201	High Impedance		
					CT2750	Implantable cardiac electrical stimulators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070101	Failure to Capture		
									A072201	High Impedance		
							F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A070101	Failure to Capture		
									A072201	High Impedance		
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A070101	Failure to Capture		
									A072201	High Impedance		
							F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A070101	Failure to Capture		
									A072201	High Impedance		
					CT2985	Active	E2403	No Clinical Signs, Symptoms or Conditions	A070101	Failure to Capture		
									A072201	High Impedance		
							F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A070101	Failure to Capture		
									A072201	High Impedance		
					CT3001	Implantable metal	E2403	No Clinical Signs, Symptoms or Conditions	A070101	Failure to Capture		
									A072201	High Impedance		
							F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A070101	Failure to Capture		
									A072201	High Impedance		
					CT3002	Implantable synthetic polymer	E2403	No Clinical Signs, Symptoms or Conditions	A070101	Failure to Capture		
									A072201	High Impedance		
							F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A070101	Failure to Capture		
									A072201	High Impedance		
					CT3006	Implantable inorganic material	E2403	No Clinical Signs, Symptoms or Conditions	A070101	Failure to Capture		
									A072201	High Impedance		
							F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A070101	Failure to Capture		
									A072201	High Impedance		
					CT314	Electrically powered	E2403	No Clinical Signs, Symptoms or Conditions	A070101	Failure to Capture		
									A072201	High Impedance		
							F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A070101	Failure to Capture		
									A072201	High Impedance		
					CT321	Long-term surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A070101	Failure to Capture		
									A072201	High Impedance		
							F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A070101	Failure to Capture		
									A072201	High Impedance		
					CT449	Generators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070101	Failure to Capture		
									A072201	High Impedance		
							F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A070101	Failure to Capture		
									A072201	High Impedance		
					CT981	Single use	E2403	No Clinical Signs, Symptoms or Conditions	A070101	Failure to Capture		
									A072201	High Impedance		
							F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A070101	Failure to Capture		
									A072201	High Impedance		
					CT983	Surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A070101	Failure to Capture		
									A072201	High Impedance		
							F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A070101	Failure to Capture		
									A072201	High Impedance		
7638	2022	All other reportable incidents	35223	Endocardial/interventricular septal pacing lead	CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070901	Decreased Sensitivity		
									A070910	Undersensing		
									A0712	Pacing Problem		
							F26	No Health Consequences or Impact	A070901	Decreased Sensitivity		
									A070910	Undersensing		
									A0712	Pacing Problem		
7639	2022	All other reportable incidents	35223	Endocardial/interventricular septal pacing lead	CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A090801	Signal Artifact/Noise	19/01/2022	19/01/2022
							F26	No Health Consequences or Impact	A090801	Signal Artifact/Noise	19/01/2022	19/01/2022
7640	2022	All other reportable incidents	35223	Endocardial/interventricular septal pacing lead	CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A072202	Low Impedance		
							F26	No Health Consequences or Impact	A072202	Low Impedance		

7641	2022	All other reportable incidents	35223	Endocardial/interventricular septal pacing lead	CT1684	Cardiac pacemakers and associated devices	E011903	Syncope/Fainting	A07	Electrical /Electronic Property Problem		
							E2403	No Clinical Signs, Symptoms or Conditions	A07	Electrical /Electronic Property Problem		
							F1905	Device Revision or Replacement	A07	Electrical /Electronic Property Problem		
7642	2022	All other reportable incidents	35223	Endocardial/interventricular septal pacing lead	CT1684	Cardiac pacemakers and associated devices	E0112	Dizziness	A070909	Oversensing		
									A23	Use of Device Problem		
							E2403	No Clinical Signs, Symptoms or Conditions	A070909	Oversensing		
									A23	Use of Device Problem		
							F23	Unexpected Medical Intervention	A070909	Oversensing		
									A23	Use of Device Problem		
7643	2022	All other reportable incidents	47265	Dual-chamber implantable pacemaker, rate-responsive	CT141	Electrodes/Leads and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT145	Cardiovascular devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT233	Surgical	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT2406	Implantable	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT2407	Active implantable	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT2727	Active implant devices/ancillaries	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT2750	Implantable cardiac electrical stimulators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT2985	Active	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT3001	Implantable metal	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT3002	Implantable synthetic polymer	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT3006	Implantable inorganic material	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT314	Electrically powered	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT321	Long-term surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT449	Generators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT981	Single use	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT983	Surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
7644	2022	All other reportable incidents	17882	Physiologic-monitoring defibrillation system	CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070803	Failure to Power Up		
							F27	Problem Identified During Non-Clinical Procedure	A070803	Failure to Power Up		
7645	2022	All other reportable incidents	47270	Cardiac resynchronization therapy implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A070910	Undersensing		
									A071302	Inappropriate/Inadequate Shock/Stimulation		
							F23	Unexpected Medical Intervention	A070910	Undersensing		
									A071302	Inappropriate/Inadequate Shock/Stimulation		
7646	2022	All other reportable incidents	47265	Dual-chamber implantable pacemaker, rate-responsive	CT141	Electrodes/Leads and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070101	Failure to Capture		
							F26	No Health Consequences or Impact	A070101	Failure to Capture		
					CT145	Cardiovascular devices	E2403	No Clinical Signs, Symptoms or Conditions	A070101	Failure to Capture		
							F26	No Health Consequences or Impact	A070101	Failure to Capture		
					CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070101	Failure to Capture		
							F26	No Health Consequences or Impact	A070101	Failure to Capture		
					CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A070101	Failure to Capture		
							F26	No Health Consequences or Impact	A070101	Failure to Capture		
					CT233	Surgical	E2403	No Clinical Signs, Symptoms or Conditions	A070101	Failure to Capture		
							F26	No Health Consequences or Impact	A070101	Failure to Capture		
					CT2406	Implantable	E2403	No Clinical Signs, Symptoms or Conditions	A070101	Failure to Capture		
							F26	No Health Consequences or Impact	A070101	Failure to Capture		
					CT2407	Active implantable	E2403	No Clinical Signs, Symptoms or Conditions	A070101	Failure to Capture		
							F26	No Health Consequences or Impact	A070101	Failure to Capture		
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A070101	Failure to Capture		
							F26	No Health Consequences or Impact	A070101	Failure to Capture		
					CT2727	Active implant devices/ancillaries	E2403	No Clinical Signs, Symptoms or Conditions	A070101	Failure to Capture		
							F26	No Health Consequences or Impact	A070101	Failure to Capture		
					CT2750	Implantable cardiac electrical stimulators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070101	Failure to Capture		
							F26	No Health Consequences or Impact	A070101	Failure to Capture		
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A070101	Failure to Capture		
							F26	No Health Consequences or Impact	A070101	Failure to Capture		
					CT2985	Active	E2403	No Clinical Signs, Symptoms or Conditions	A070101	Failure to Capture		
							F26	No Health Consequences or Impact	A070101	Failure to Capture		
					CT3001	Implantable metal	E2403	No Clinical Signs, Symptoms or Conditions	A070101	Failure to Capture		
							F26	No Health Consequences or Impact	A070101	Failure to Capture		
					CT3002	Implantable synthetic polymer	E2403	No Clinical Signs, Symptoms or Conditions	A070101	Failure to Capture		
							F26	No Health Consequences or Impact	A070101	Failure to Capture		
					CT3006	Implantable inorganic material	E2403	No Clinical Signs, Symptoms or Conditions	A070101	Failure to Capture		
							F26	No Health Consequences or Impact	A070101	Failure to Capture		
					CT314	Electrically powered	E2403	No Clinical Signs, Symptoms or Conditions	A070101	Failure to Capture		
							F26	No Health Consequences or Impact	A070101	Failure to Capture		
					CT321	Long-term surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A070101	Failure to Capture		

						F26	No Health Consequences or Impact	A070101	Failure to Capture			
					CT449	Generators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070101	Failure to Capture		
							F26	No Health Consequences or Impact	A070101	Failure to Capture		
					CT981	Single use	E2403	No Clinical Signs, Symptoms or Conditions	A070101	Failure to Capture		
							F26	No Health Consequences or Impact	A070101	Failure to Capture		
					CT983	Surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A070101	Failure to Capture		
							F26	No Health Consequences or Impact	A070101	Failure to Capture		
7647	2022	All other reportable incidents	17882	Physiologic-monitoring defibrillation system	CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A0801	Failure to Calibrate		
							F27	Problem Identified During Non-Clinical Procedure	A0801	Failure to Calibrate		
7648	2022	All other reportable incidents	17882	Physiologic-monitoring defibrillation system	CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A1603	Failure of Device to Self-Test		
							F27	Problem Identified During Non-Clinical Procedure	A1603	Failure of Device to Self-Test		
7649	2022	All other reportable incidents	35852	Single-chamber implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem	01/01/2022	31/12/2022
							F08	Hospitalization or Prolonged Hospitalization	A0705	Battery Problem	01/01/2022	31/12/2022
							F12	Serious Injury/ Illness/ Impairment	A0705	Battery Problem	01/01/2022	31/12/2022
							F1905	Device Revision or Replacement	A0705	Battery Problem	01/01/2022	31/12/2022
7650	2022	All other reportable incidents	47265	Dual-chamber implantable pacemaker, rate-responsive	CT141	Electrodes/Leads and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem		
									A150103	Premature Activation		
							F26	No Health Consequences or Impact	A0705	Battery Problem		
									A150103	Premature Activation		
					CT145	Cardiovascular devices	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem		
									A150103	Premature Activation		
							F26	No Health Consequences or Impact	A0705	Battery Problem		
									A150103	Premature Activation		
					CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem		
									A150103	Premature Activation		
							F26	No Health Consequences or Impact	A0705	Battery Problem		
									A150103	Premature Activation		
					CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem		
									A150103	Premature Activation		
							F26	No Health Consequences or Impact	A0705	Battery Problem		
									A150103	Premature Activation		
					CT233	Surgical	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem		
									A150103	Premature Activation		
							F26	No Health Consequences or Impact	A0705	Battery Problem		
									A150103	Premature Activation		
					CT2406	Implantable	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem		
									A150103	Premature Activation		
							F26	No Health Consequences or Impact	A0705	Battery Problem		
									A150103	Premature Activation		
					CT2407	Active implantable	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem		
									A150103	Premature Activation		
							F26	No Health Consequences or Impact	A0705	Battery Problem		
									A150103	Premature Activation		
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem		
									A150103	Premature Activation		
							F26	No Health Consequences or Impact	A0705	Battery Problem		
									A150103	Premature Activation		
					CT2727	Active implant devices/ancillaries	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem		
									A150103	Premature Activation		
							F26	No Health Consequences or Impact	A0705	Battery Problem		
									A150103	Premature Activation		
					CT2750	Implantable cardiac electrical stimulators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem		
									A150103	Premature Activation		
							F26	No Health Consequences or Impact	A0705	Battery Problem		
									A150103	Premature Activation		
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem		
									A150103	Premature Activation		
							F26	No Health Consequences or Impact	A0705	Battery Problem		
									A150103	Premature Activation		
					CT2985	Active	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem		
									A150103	Premature Activation		
							F26	No Health Consequences or Impact	A0705	Battery Problem		
									A150103	Premature Activation		
					CT3001	Implantable metal	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem		
									A150103	Premature Activation		
							F26	No Health Consequences or Impact	A0705	Battery Problem		
									A150103	Premature Activation		
					CT3002	Implantable synthetic polymer	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem		
									A150103	Premature Activation		
							F26	No Health Consequences or Impact	A0705	Battery Problem		
									A150103	Premature Activation		
					CT3006	Implantable inorganic material	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem		
									A150103	Premature Activation		
							F26	No Health Consequences or Impact	A0705	Battery Problem		
									A150103	Premature Activation		
					CT314	Electrically powered	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem		
									A150103	Premature Activation		
							F26	No Health Consequences or Impact	A0705	Battery Problem		
									A150103	Premature Activation		
					CT321	Long-term surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem		
									A150103	Premature Activation		
							F26	No Health Consequences or Impact	A0705	Battery Problem		
									A150103	Premature Activation		
					CT449	Generators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem		
									A150103	Premature Activation		
							F26	No Health Consequences or Impact	A0705	Battery Problem		
									A150103	Premature Activation		
					CT981	Single use	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem		

								A150103	Premature Activation		
						F26	No Health Consequences or Impact	A0705	Battery Problem		
								A150103	Premature Activation		
					CT983	Surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem		
									A150103	Premature Activation		
							F26	No Health Consequences or Impact	A0705	Battery Problem		
									A150103	Premature Activation		
7651	2022	All other reportable incidents	47263	Cardiac resynchronization therapy implantable pacemaker	CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A0908	Incorrect, Inadequate or Imprecise Result or Readings		
							F26	No Health Consequences or Impact	A0908	Incorrect, Inadequate or Imprecise Result or Readings		
					CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A0908	Incorrect, Inadequate or Imprecise Result or Readings		
							F26	No Health Consequences or Impact	A0908	Incorrect, Inadequate or Imprecise Result or Readings		
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A0908	Incorrect, Inadequate or Imprecise Result or Readings		
							F26	No Health Consequences or Impact	A0908	Incorrect, Inadequate or Imprecise Result or Readings		
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A0908	Incorrect, Inadequate or Imprecise Result or Readings		
							F26	No Health Consequences or Impact	A0908	Incorrect, Inadequate or Imprecise Result or Readings		
7652	2022	All other reportable incidents	36241	Pacemaker/defibrillator lead	CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070101	Failure to Capture	18/09/2021	18/09/2021
									A070102	High Capture Threshold	18/09/2021	18/09/2021
							F08	Hospitalization or Prolonged Hospitalization	A070101	Failure to Capture	18/09/2021	18/09/2021
									A070102	High Capture Threshold	18/09/2021	18/09/2021
							F1905	Device Revision or Replacement	A070101	Failure to Capture	18/09/2021	18/09/2021
									A070102	High Capture Threshold	18/09/2021	18/09/2021
7653	2022	All other reportable incidents	35223	Endocardial/interventricular septal pacing lead	CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A0401	Break		
									A070909	Oversensing		
									A0722	Impedance Problem		
							F26	No Health Consequences or Impact	A0401	Break		
									A070909	Oversensing		
									A0722	Impedance Problem		
7654	2022	All other reportable incidents	36241	Pacemaker/defibrillator lead	CT1684	Cardiac pacemakers and associated devices	E2104	Electric Shock	A0508	Noise, Audible		
									A070909	Oversensing		
									A071302	Inappropriate/Inadequate Shock/Stimulation		
							E2403	No Clinical Signs, Symptoms or Conditions	A0508	Noise, Audible		
									A070909	Oversensing		
									A071302	Inappropriate/Inadequate Shock/Stimulation		
							F08	Hospitalization or Prolonged Hospitalization	A0508	Noise, Audible		
									A070909	Oversensing		
									A071302	Inappropriate/Inadequate Shock/Stimulation		
							F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A0508	Noise, Audible		
									A070909	Oversensing		
									A071302	Inappropriate/Inadequate Shock/Stimulation		
							F1905	Device Revision or Replacement	A0508	Noise, Audible		
									A070909	Oversensing		
									A071302	Inappropriate/Inadequate Shock/Stimulation		
7655	2022	All other reportable incidents	47267	Single-chamber implantable pacemaker, rate-responsive	CT141	Electrodes/Leads and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	28/10/2021	28/10/2021
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	28/10/2021	28/10/2021
							F1903	Device Explantation	A070504	Premature Discharge of Battery	28/10/2021	28/10/2021
					CT145	Cardiovascular devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	28/10/2021	28/10/2021
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	28/10/2021	28/10/2021
							F1903	Device Explantation	A070504	Premature Discharge of Battery	28/10/2021	28/10/2021
					CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	28/10/2021	28/10/2021
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	28/10/2021	28/10/2021
							F1903	Device Explantation	A070504	Premature Discharge of Battery	28/10/2021	28/10/2021
					CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	28/10/2021	28/10/2021
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	28/10/2021	28/10/2021
							F1903	Device Explantation	A070504	Premature Discharge of Battery	28/10/2021	28/10/2021
					CT233	Surgical	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	28/10/2021	28/10/2021
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	28/10/2021	28/10/2021
							F1903	Device Explantation	A070504	Premature Discharge of Battery	28/10/2021	28/10/2021
					CT2406	Implantable	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	28/10/2021	28/10/2021
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	28/10/2021	28/10/2021
							F1903	Device Explantation	A070504	Premature Discharge of Battery	28/10/2021	28/10/2021
					CT2407	Active implantable	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	28/10/2021	28/10/2021
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	28/10/2021	28/10/2021
							F1903	Device Explantation	A070504	Premature Discharge of Battery	28/10/2021	28/10/2021
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	28/10/2021	28/10/2021
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	28/10/2021	28/10/2021
							F1903	Device Explantation	A070504	Premature Discharge of Battery	28/10/2021	28/10/2021
					CT2727	Active implant devices/ancillaries	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	28/10/2021	28/10/2021
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	28/10/2021	28/10/2021
							F1903	Device Explantation	A070504	Premature Discharge of Battery	28/10/2021	28/10/2021
					CT2750	Implantable cardiac electrical stimulators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	28/10/2021	28/10/2021
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	28/10/2021	28/10/2021
							F1903	Device Explantation	A070504	Premature Discharge of Battery	28/10/2021	28/10/2021
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	28/10/2021	28/10/2021
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	28/10/2021	28/10/2021
							F1903	Device Explantation	A070504	Premature Discharge of Battery	28/10/2021	28/10/2021
					CT2985	Active	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	28/10/2021	28/10/2021
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	28/10/2021	28/10/2021
							F1903	Device Explantation	A070504	Premature Discharge of Battery	28/10/2021	28/10/2021
					CT3001	Implantable metal	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	28/10/2021	28/10/2021
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	28/10/2021	28/10/2021
							F1903	Device Explantation	A070504	Premature Discharge of Battery	28/10/2021	28/10/2021
					CT3002	Implantable synthetic polymer	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	28/10/2021	28/10/2021

							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	28/10/2021	28/10/2021
							F1903	Device Explantation	A070504	Premature Discharge of Battery	28/10/2021	28/10/2021
					CT3006	Implantable inorganic material	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	28/10/2021	28/10/2021
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	28/10/2021	28/10/2021
							F1903	Device Explantation	A070504	Premature Discharge of Battery	28/10/2021	28/10/2021
					CT314	Electrically powered	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	28/10/2021	28/10/2021
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	28/10/2021	28/10/2021
							F1903	Device Explantation	A070504	Premature Discharge of Battery	28/10/2021	28/10/2021
					CT321	Long-term surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	28/10/2021	28/10/2021
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	28/10/2021	28/10/2021
							F1903	Device Explantation	A070504	Premature Discharge of Battery	28/10/2021	28/10/2021
					CT449	Generators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	28/10/2021	28/10/2021
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	28/10/2021	28/10/2021
							F1903	Device Explantation	A070504	Premature Discharge of Battery	28/10/2021	28/10/2021
					CT981	Single use	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	28/10/2021	28/10/2021
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	28/10/2021	28/10/2021
							F1903	Device Explantation	A070504	Premature Discharge of Battery	28/10/2021	28/10/2021
					CT983	Surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	28/10/2021	28/10/2021
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	28/10/2021	28/10/2021
							F1903	Device Explantation	A070504	Premature Discharge of Battery	28/10/2021	28/10/2021
7656	2022	All other reportable incidents	35852	Single-chamber implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	13/01/2022	13/01/2022
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	13/01/2022	13/01/2022
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	13/01/2022	13/01/2022
7657	2022	All other reportable incidents	35852	Single-chamber implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	03/02/2022	03/02/2022
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	03/02/2022	03/02/2022
							F1903	Device Explantation	A070504	Premature Discharge of Battery	03/02/2022	03/02/2022
7658	2022	All other reportable incidents	17882	Physiologic-monitoring defibrillation system	CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070803	Failure to Power Up		
							F27	Problem Identified During Non-Clinical Procedure	A070803	Failure to Power Up		
7659	2022	All other reportable incidents	35852	Single-chamber implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem	09/09/2022	09/09/2022
									A070504	Premature Discharge of Battery	09/09/2022	09/09/2022
							F08	Hospitalization or Prolonged Hospitalization	A0705	Battery Problem	09/09/2022	09/09/2022
									A070504	Premature Discharge of Battery	09/09/2022	09/09/2022
							F1905	Device Revision or Replacement	A0705	Battery Problem	09/09/2022	09/09/2022
									A070504	Premature Discharge of Battery	09/09/2022	09/09/2022
7660	2022	All other reportable incidents	47263	Cardiac resynchronization therapy implantable pacemaker	CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A0908	Incorrect, Inadequate or Imprecise Result or Readings		
							F26	No Health Consequences or Impact	A0908	Incorrect, Inadequate or Imprecise Result or Readings		
					CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A0908	Incorrect, Inadequate or Imprecise Result or Readings		
							F26	No Health Consequences or Impact	A0908	Incorrect, Inadequate or Imprecise Result or Readings		
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A0908	Incorrect, Inadequate or Imprecise Result or Readings		
							F26	No Health Consequences or Impact	A0908	Incorrect, Inadequate or Imprecise Result or Readings		
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A0908	Incorrect, Inadequate or Imprecise Result or Readings		
							F26	No Health Consequences or Impact	A0908	Incorrect, Inadequate or Imprecise Result or Readings		
7661	2022	All other reportable incidents	17882	Physiologic-monitoring defibrillation system	CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A07	Electrical /Electronic Property Problem		
							F27	Problem Identified During Non-Clinical Procedure	A07	Electrical /Electronic Property Problem		
7662	2022	All other reportable incidents	17882	Physiologic-monitoring defibrillation system	CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A07	Electrical /Electronic Property Problem		
							F27	Problem Identified During Non-Clinical Procedure	A07	Electrical /Electronic Property Problem		
7663	2022	All other reportable incidents	17882	Physiologic-monitoring defibrillation system	CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A07	Electrical /Electronic Property Problem		
							F27	Problem Identified During Non-Clinical Procedure	A07	Electrical /Electronic Property Problem		
7664	2022	All other reportable incidents	17882	Physiologic-monitoring defibrillation system	CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A07	Electrical /Electronic Property Problem		
							F27	Problem Identified During Non-Clinical Procedure	A07	Electrical /Electronic Property Problem		
7665	2022	All other reportable incidents	17882	Physiologic-monitoring defibrillation system	CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A07	Electrical /Electronic Property Problem		
							F27	Problem Identified During Non-Clinical Procedure	A07	Electrical /Electronic Property Problem		
7666	2022	All other reportable incidents	17882	Physiologic-monitoring defibrillation system	CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A07	Electrical /Electronic Property Problem		
							F27	Problem Identified During Non-Clinical Procedure	A07	Electrical /Electronic Property Problem		
7667	2022	All other reportable incidents	47265	Dual-chamber implantable pacemaker, rate-responsive	CT141	Electrodes/Leads and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F1901	Additional Surgery	A070504	Premature Discharge of Battery		
					CT145	Cardiovascular devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F1901	Additional Surgery	A070504	Premature Discharge of Battery		
					CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F1901	Additional Surgery	A070504	Premature Discharge of Battery		
					CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F1901	Additional Surgery	A070504	Premature Discharge of Battery		
					CT233	Surgical	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F1901	Additional Surgery	A070504	Premature Discharge of Battery		
					CT2406	Implantable	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F1901	Additional Surgery	A070504	Premature Discharge of Battery		
					CT2407	Active implantable	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F1901	Additional Surgery	A070504	Premature Discharge of Battery		
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F1901	Additional Surgery	A070504	Premature Discharge of Battery		
					CT2727	Active implant devices/ancillaries	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F1901	Additional Surgery	A070504	Premature Discharge of Battery		
					CT2750	Implantable cardiac electrical stimulators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F1901	Additional Surgery	A070504	Premature Discharge of Battery		
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F1901	Additional Surgery	A070504	Premature Discharge of Battery		
					CT2985	Active	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F1901	Additional Surgery	A070504	Premature Discharge of Battery		
					CT3001	Implantable metal	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F1901	Additional Surgery	A070504	Premature Discharge of Battery		
					CT3002	Implantable synthetic polymer	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F1901	Additional Surgery	A070504	Premature Discharge of Battery		

					CT3006	Implantable inorganic material	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F1901	Additional Surgery	A070504	Premature Discharge of Battery		
					CT314	Electrically powered	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F1901	Additional Surgery	A070504	Premature Discharge of Battery		
					CT321	Long-term surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F1901	Additional Surgery	A070504	Premature Discharge of Battery		
					CT449	Generators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F1901	Additional Surgery	A070504	Premature Discharge of Battery		
					CT981	Single use	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F1901	Additional Surgery	A070504	Premature Discharge of Battery		
					CT983	Surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F1901	Additional Surgery	A070504	Premature Discharge of Battery		
7668	2022	All other reportable incidents	17882	Physiologic-monitoring defibrillation system	CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A07	Electrical /Electronic Property Problem		
							F27	Problem Identified During Non-Clinical Procedure	A07	Electrical /Electronic Property Problem		
7669	2022	All other reportable incidents	35852	Single-chamber implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem		
							F12	Serious Injury/ Illness/ Impairment	A0705	Battery Problem		
							F23	Unexpected Medical Intervention	A0705	Battery Problem		
7670	2022	All other reportable incidents	47265	Dual-chamber implantable pacemaker, rate-responsive	CT141	Electrodes/Leads and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem	11/03/2022	11/03/2022
									A070504	Premature Discharge of Battery	11/03/2022	11/03/2022
							F08	Hospitalization or Prolonged Hospitalization	A0705	Battery Problem	11/03/2022	11/03/2022
									A070504	Premature Discharge of Battery	11/03/2022	11/03/2022
							F1905	Device Revision or Replacement	A0705	Battery Problem	11/03/2022	11/03/2022
									A070504	Premature Discharge of Battery	11/03/2022	11/03/2022
					CT145	Cardiovascular devices	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem	11/03/2022	11/03/2022
									A070504	Premature Discharge of Battery	11/03/2022	11/03/2022
							F08	Hospitalization or Prolonged Hospitalization	A0705	Battery Problem	11/03/2022	11/03/2022
									A070504	Premature Discharge of Battery	11/03/2022	11/03/2022
							F1905	Device Revision or Replacement	A0705	Battery Problem	11/03/2022	11/03/2022
									A070504	Premature Discharge of Battery	11/03/2022	11/03/2022
					CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem	11/03/2022	11/03/2022
									A070504	Premature Discharge of Battery	11/03/2022	11/03/2022
							F08	Hospitalization or Prolonged Hospitalization	A0705	Battery Problem	11/03/2022	11/03/2022
									A070504	Premature Discharge of Battery	11/03/2022	11/03/2022
							F1905	Device Revision or Replacement	A0705	Battery Problem	11/03/2022	11/03/2022
									A070504	Premature Discharge of Battery	11/03/2022	11/03/2022
					CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem	11/03/2022	11/03/2022
									A070504	Premature Discharge of Battery	11/03/2022	11/03/2022
							F08	Hospitalization or Prolonged Hospitalization	A0705	Battery Problem	11/03/2022	11/03/2022
									A070504	Premature Discharge of Battery	11/03/2022	11/03/2022
							F1905	Device Revision or Replacement	A0705	Battery Problem	11/03/2022	11/03/2022
									A070504	Premature Discharge of Battery	11/03/2022	11/03/2022
					CT233	Surgical	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem	11/03/2022	11/03/2022
									A070504	Premature Discharge of Battery	11/03/2022	11/03/2022
							F08	Hospitalization or Prolonged Hospitalization	A0705	Battery Problem	11/03/2022	11/03/2022
									A070504	Premature Discharge of Battery	11/03/2022	11/03/2022
							F1905	Device Revision or Replacement	A0705	Battery Problem	11/03/2022	11/03/2022
									A070504	Premature Discharge of Battery	11/03/2022	11/03/2022
					CT2406	Implantable	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem	11/03/2022	11/03/2022
									A070504	Premature Discharge of Battery	11/03/2022	11/03/2022
							F08	Hospitalization or Prolonged Hospitalization	A0705	Battery Problem	11/03/2022	11/03/2022
									A070504	Premature Discharge of Battery	11/03/2022	11/03/2022
							F1905	Device Revision or Replacement	A0705	Battery Problem	11/03/2022	11/03/2022
									A070504	Premature Discharge of Battery	11/03/2022	11/03/2022
					CT2407	Active implantable	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem	11/03/2022	11/03/2022
									A070504	Premature Discharge of Battery	11/03/2022	11/03/2022
							F08	Hospitalization or Prolonged Hospitalization	A0705	Battery Problem	11/03/2022	11/03/2022
									A070504	Premature Discharge of Battery	11/03/2022	11/03/2022
							F1905	Device Revision or Replacement	A0705	Battery Problem	11/03/2022	11/03/2022
									A070504	Premature Discharge of Battery	11/03/2022	11/03/2022
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem	11/03/2022	11/03/2022
									A070504	Premature Discharge of Battery	11/03/2022	11/03/2022
							F08	Hospitalization or Prolonged Hospitalization	A0705	Battery Problem	11/03/2022	11/03/2022
									A070504	Premature Discharge of Battery	11/03/2022	11/03/2022
							F1905	Device Revision or Replacement	A0705	Battery Problem	11/03/2022	11/03/2022
									A070504	Premature Discharge of Battery	11/03/2022	11/03/2022
					CT2727	Active implant devices/ancillaries	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem	11/03/2022	11/03/2022
									A070504	Premature Discharge of Battery	11/03/2022	11/03/2022
							F08	Hospitalization or Prolonged Hospitalization	A0705	Battery Problem	11/03/2022	11/03/2022
									A070504	Premature Discharge of Battery	11/03/2022	11/03/2022
							F1905	Device Revision or Replacement	A0705	Battery Problem	11/03/2022	11/03/2022
									A070504	Premature Discharge of Battery	11/03/2022	11/03/2022
					CT2750	Implantable cardiac electrical stimulators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem	11/03/2022	11/03/2022
									A070504	Premature Discharge of Battery	11/03/2022	11/03/2022
							F08	Hospitalization or Prolonged Hospitalization	A0705	Battery Problem	11/03/2022	11/03/2022
									A070504	Premature Discharge of Battery	11/03/2022	11/03/2022
							F1905	Device Revision or Replacement	A0705	Battery Problem	11/03/2022	11/03/2022
									A070504	Premature Discharge of Battery	11/03/2022	11/03/2022
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem	11/03/2022	11/03/2022
									A070504	Premature Discharge of Battery	11/03/2022	11/03/2022
							F08	Hospitalization or Prolonged Hospitalization	A0705	Battery Problem	11/03/2022	11/03/2022
									A070504	Premature Discharge of Battery	11/03/2022	11/03/2022
							F1905	Device Revision or Replacement	A0705	Battery Problem	11/03/2022	11/03/2022
									A070504	Premature Discharge of Battery	11/03/2022	11/03/2022
					CT2985	Active	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem	11/03/2022	11/03/2022
									A070504	Premature Discharge of Battery	11/03/2022	11/03/2022
							F08	Hospitalization or Prolonged Hospitalization	A0705	Battery Problem	11/03/2022	11/03/2022
									A070504	Premature Discharge of Battery	11/03/2022	11/03/2022
							F1905	Device Revision or Replacement	A0705	Battery Problem	11/03/2022	11/03/2022

									A070504	Premature Discharge of Battery	11/03/2022	11/03/2022
					CT3001	Implantable metal	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem	11/03/2022	11/03/2022
									A070504	Premature Discharge of Battery	11/03/2022	11/03/2022
							F08	Hospitalization or Prolonged Hospitalization	A0705	Battery Problem	11/03/2022	11/03/2022
									A070504	Premature Discharge of Battery	11/03/2022	11/03/2022
							F1905	Device Revision or Replacement	A0705	Battery Problem	11/03/2022	11/03/2022
									A070504	Premature Discharge of Battery	11/03/2022	11/03/2022
					CT3002	Implantable synthetic polymer	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem	11/03/2022	11/03/2022
									A070504	Premature Discharge of Battery	11/03/2022	11/03/2022
							F08	Hospitalization or Prolonged Hospitalization	A0705	Battery Problem	11/03/2022	11/03/2022
									A070504	Premature Discharge of Battery	11/03/2022	11/03/2022
							F1905	Device Revision or Replacement	A0705	Battery Problem	11/03/2022	11/03/2022
									A070504	Premature Discharge of Battery	11/03/2022	11/03/2022
					CT3006	Implantable inorganic material	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem	11/03/2022	11/03/2022
									A070504	Premature Discharge of Battery	11/03/2022	11/03/2022
							F08	Hospitalization or Prolonged Hospitalization	A0705	Battery Problem	11/03/2022	11/03/2022
									A070504	Premature Discharge of Battery	11/03/2022	11/03/2022
							F1905	Device Revision or Replacement	A0705	Battery Problem	11/03/2022	11/03/2022
									A070504	Premature Discharge of Battery	11/03/2022	11/03/2022
					CT314	Electrically powered	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem	11/03/2022	11/03/2022
									A070504	Premature Discharge of Battery	11/03/2022	11/03/2022
							F08	Hospitalization or Prolonged Hospitalization	A0705	Battery Problem	11/03/2022	11/03/2022
									A070504	Premature Discharge of Battery	11/03/2022	11/03/2022
							F1905	Device Revision or Replacement	A0705	Battery Problem	11/03/2022	11/03/2022
									A070504	Premature Discharge of Battery	11/03/2022	11/03/2022
					CT321	Long-term surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem	11/03/2022	11/03/2022
									A070504	Premature Discharge of Battery	11/03/2022	11/03/2022
							F08	Hospitalization or Prolonged Hospitalization	A0705	Battery Problem	11/03/2022	11/03/2022
									A070504	Premature Discharge of Battery	11/03/2022	11/03/2022
							F1905	Device Revision or Replacement	A0705	Battery Problem	11/03/2022	11/03/2022
									A070504	Premature Discharge of Battery	11/03/2022	11/03/2022
					CT449	Generators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem	11/03/2022	11/03/2022
									A070504	Premature Discharge of Battery	11/03/2022	11/03/2022
							F08	Hospitalization or Prolonged Hospitalization	A0705	Battery Problem	11/03/2022	11/03/2022
									A070504	Premature Discharge of Battery	11/03/2022	11/03/2022
							F1905	Device Revision or Replacement	A0705	Battery Problem	11/03/2022	11/03/2022
									A070504	Premature Discharge of Battery	11/03/2022	11/03/2022
					CT981	Single use	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem	11/03/2022	11/03/2022
									A070504	Premature Discharge of Battery	11/03/2022	11/03/2022
							F08	Hospitalization or Prolonged Hospitalization	A0705	Battery Problem	11/03/2022	11/03/2022
									A070504	Premature Discharge of Battery	11/03/2022	11/03/2022
							F1905	Device Revision or Replacement	A0705	Battery Problem	11/03/2022	11/03/2022
									A070504	Premature Discharge of Battery	11/03/2022	11/03/2022
					CT983	Surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem	11/03/2022	11/03/2022
									A070504	Premature Discharge of Battery	11/03/2022	11/03/2022
							F08	Hospitalization or Prolonged Hospitalization	A0705	Battery Problem	11/03/2022	11/03/2022
									A070504	Premature Discharge of Battery	11/03/2022	11/03/2022
							F1905	Device Revision or Replacement	A0705	Battery Problem	11/03/2022	11/03/2022
									A070504	Premature Discharge of Battery	11/03/2022	11/03/2022
7671	2022	All other reportable incidents	47265	Dual-chamber implantable pacemaker, rate-responsive	CT141	Electrodes/Leads and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem	02/03/2022	02/03/2022
									A070504	Premature Discharge of Battery	02/03/2022	02/03/2022
							F08	Hospitalization or Prolonged Hospitalization	A0705	Battery Problem	02/03/2022	02/03/2022
									A070504	Premature Discharge of Battery	02/03/2022	02/03/2022
							F1905	Device Revision or Replacement	A0705	Battery Problem	02/03/2022	02/03/2022
									A070504	Premature Discharge of Battery	02/03/2022	02/03/2022
					CT145	Cardiovascular devices	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem	02/03/2022	02/03/2022
									A070504	Premature Discharge of Battery	02/03/2022	02/03/2022
							F08	Hospitalization or Prolonged Hospitalization	A0705	Battery Problem	02/03/2022	02/03/2022
									A070504	Premature Discharge of Battery	02/03/2022	02/03/2022
							F1905	Device Revision or Replacement	A0705	Battery Problem	02/03/2022	02/03/2022
									A070504	Premature Discharge of Battery	02/03/2022	02/03/2022
					CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem	02/03/2022	02/03/2022
									A070504	Premature Discharge of Battery	02/03/2022	02/03/2022
							F08	Hospitalization or Prolonged Hospitalization	A0705	Battery Problem	02/03/2022	02/03/2022
									A070504	Premature Discharge of Battery	02/03/2022	02/03/2022
							F1905	Device Revision or Replacement	A0705	Battery Problem	02/03/2022	02/03/2022
									A070504	Premature Discharge of Battery	02/03/2022	02/03/2022
					CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem	02/03/2022	02/03/2022
									A070504	Premature Discharge of Battery	02/03/2022	02/03/2022
							F08	Hospitalization or Prolonged Hospitalization	A0705	Battery Problem	02/03/2022	02/03/2022
									A070504	Premature Discharge of Battery	02/03/2022	02/03/2022
							F1905	Device Revision or Replacement	A0705	Battery Problem	02/03/2022	02/03/2022
									A070504	Premature Discharge of Battery	02/03/2022	02/03/2022
					CT233	Surgical	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem	02/03/2022	02/03/2022
									A070504	Premature Discharge of Battery	02/03/2022	02/03/2022
							F08	Hospitalization or Prolonged Hospitalization	A0705	Battery Problem	02/03/2022	02/03/2022
									A070504	Premature Discharge of Battery	02/03/2022	02/03/2022
							F1905	Device Revision or Replacement	A0705	Battery Problem	02/03/2022	02/03/2022
									A070504	Premature Discharge of Battery	02/03/2022	02/03/2022
					CT2406	Implantable	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem	02/03/2022	02/03/2022
									A070504	Premature Discharge of Battery	02/03/2022	02/03/2022
							F08	Hospitalization or Prolonged Hospitalization	A0705	Battery Problem	02/03/2022	02/03/2022
									A070504	Premature Discharge of Battery	02/03/2022	02/03/2022
							F1905	Device Revision or Replacement	A0705	Battery Problem	02/03/2022	02/03/2022
									A070504	Premature Discharge of Battery	02/03/2022	02/03/2022
					CT2407	Active implantable	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem	02/03/2022	02/03/2022
									A070504	Premature Discharge of Battery	02/03/2022	02/03/2022
							F08	Hospitalization or Prolonged Hospitalization	A0705	Battery Problem	02/03/2022	02/03/2022

								A070504	Premature Discharge of Battery	02/03/2022	02/03/2022
						F1905	Device Revision or Replacement	A0705	Battery Problem	02/03/2022	02/03/2022
								A070504	Premature Discharge of Battery	02/03/2022	02/03/2022
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem	02/03/2022	02/03/2022
									A070504	Premature Discharge of Battery	02/03/2022	02/03/2022
							F08	Hospitalization or Prolonged Hospitalization	A0705	Battery Problem	02/03/2022	02/03/2022
									A070504	Premature Discharge of Battery	02/03/2022	02/03/2022
							F1905	Device Revision or Replacement	A0705	Battery Problem	02/03/2022	02/03/2022
									A070504	Premature Discharge of Battery	02/03/2022	02/03/2022
					CT2727	Active implant devices/ancillaries	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem	02/03/2022	02/03/2022
									A070504	Premature Discharge of Battery	02/03/2022	02/03/2022
							F08	Hospitalization or Prolonged Hospitalization	A0705	Battery Problem	02/03/2022	02/03/2022
									A070504	Premature Discharge of Battery	02/03/2022	02/03/2022
							F1905	Device Revision or Replacement	A0705	Battery Problem	02/03/2022	02/03/2022
									A070504	Premature Discharge of Battery	02/03/2022	02/03/2022
					CT2750	Implantable cardiac electrical stimulators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem	02/03/2022	02/03/2022
									A070504	Premature Discharge of Battery	02/03/2022	02/03/2022
							F08	Hospitalization or Prolonged Hospitalization	A0705	Battery Problem	02/03/2022	02/03/2022
									A070504	Premature Discharge of Battery	02/03/2022	02/03/2022
							F1905	Device Revision or Replacement	A0705	Battery Problem	02/03/2022	02/03/2022
									A070504	Premature Discharge of Battery	02/03/2022	02/03/2022
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem	02/03/2022	02/03/2022
									A070504	Premature Discharge of Battery	02/03/2022	02/03/2022
							F08	Hospitalization or Prolonged Hospitalization	A0705	Battery Problem	02/03/2022	02/03/2022
									A070504	Premature Discharge of Battery	02/03/2022	02/03/2022
							F1905	Device Revision or Replacement	A0705	Battery Problem	02/03/2022	02/03/2022
									A070504	Premature Discharge of Battery	02/03/2022	02/03/2022
					CT2985	Active	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem	02/03/2022	02/03/2022
									A070504	Premature Discharge of Battery	02/03/2022	02/03/2022
							F08	Hospitalization or Prolonged Hospitalization	A0705	Battery Problem	02/03/2022	02/03/2022
									A070504	Premature Discharge of Battery	02/03/2022	02/03/2022
							F1905	Device Revision or Replacement	A0705	Battery Problem	02/03/2022	02/03/2022
									A070504	Premature Discharge of Battery	02/03/2022	02/03/2022
					CT3001	Implantable metal	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem	02/03/2022	02/03/2022
									A070504	Premature Discharge of Battery	02/03/2022	02/03/2022
							F08	Hospitalization or Prolonged Hospitalization	A0705	Battery Problem	02/03/2022	02/03/2022
									A070504	Premature Discharge of Battery	02/03/2022	02/03/2022
							F1905	Device Revision or Replacement	A0705	Battery Problem	02/03/2022	02/03/2022
									A070504	Premature Discharge of Battery	02/03/2022	02/03/2022
					CT3002	Implantable synthetic polymer	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem	02/03/2022	02/03/2022
									A070504	Premature Discharge of Battery	02/03/2022	02/03/2022
							F08	Hospitalization or Prolonged Hospitalization	A0705	Battery Problem	02/03/2022	02/03/2022
									A070504	Premature Discharge of Battery	02/03/2022	02/03/2022
							F1905	Device Revision or Replacement	A0705	Battery Problem	02/03/2022	02/03/2022
									A070504	Premature Discharge of Battery	02/03/2022	02/03/2022
					CT3006	Implantable inorganic material	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem	02/03/2022	02/03/2022
									A070504	Premature Discharge of Battery	02/03/2022	02/03/2022
							F08	Hospitalization or Prolonged Hospitalization	A0705	Battery Problem	02/03/2022	02/03/2022
									A070504	Premature Discharge of Battery	02/03/2022	02/03/2022
							F1905	Device Revision or Replacement	A0705	Battery Problem	02/03/2022	02/03/2022
									A070504	Premature Discharge of Battery	02/03/2022	02/03/2022
					CT314	Electrically powered	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem	02/03/2022	02/03/2022
									A070504	Premature Discharge of Battery	02/03/2022	02/03/2022
							F08	Hospitalization or Prolonged Hospitalization	A0705	Battery Problem	02/03/2022	02/03/2022
									A070504	Premature Discharge of Battery	02/03/2022	02/03/2022
							F1905	Device Revision or Replacement	A0705	Battery Problem	02/03/2022	02/03/2022
									A070504	Premature Discharge of Battery	02/03/2022	02/03/2022
					CT321	Long-term surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem	02/03/2022	02/03/2022
									A070504	Premature Discharge of Battery	02/03/2022	02/03/2022
							F08	Hospitalization or Prolonged Hospitalization	A0705	Battery Problem	02/03/2022	02/03/2022
									A070504	Premature Discharge of Battery	02/03/2022	02/03/2022
							F1905	Device Revision or Replacement	A0705	Battery Problem	02/03/2022	02/03/2022
									A070504	Premature Discharge of Battery	02/03/2022	02/03/2022
					CT449	Generators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem	02/03/2022	02/03/2022
									A070504	Premature Discharge of Battery	02/03/2022	02/03/2022
							F08	Hospitalization or Prolonged Hospitalization	A0705	Battery Problem	02/03/2022	02/03/2022
									A070504	Premature Discharge of Battery	02/03/2022	02/03/2022
							F1905	Device Revision or Replacement	A0705	Battery Problem	02/03/2022	02/03/2022
									A070504	Premature Discharge of Battery	02/03/2022	02/03/2022
					CT981	Single use	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem	02/03/2022	02/03/2022
									A070504	Premature Discharge of Battery	02/03/2022	02/03/2022
							F08	Hospitalization or Prolonged Hospitalization	A0705	Battery Problem	02/03/2022	02/03/2022
									A070504	Premature Discharge of Battery	02/03/2022	02/03/2022
							F1905	Device Revision or Replacement	A0705	Battery Problem	02/03/2022	02/03/2022
									A070504	Premature Discharge of Battery	02/03/2022	02/03/2022
					CT983	Surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem	02/03/2022	02/03/2022
									A070504	Premature Discharge of Battery	02/03/2022	02/03/2022
							F08	Hospitalization or Prolonged Hospitalization	A0705	Battery Problem	02/03/2022	02/03/2022
									A070504	Premature Discharge of Battery	02/03/2022	02/03/2022
							F1905	Device Revision or Replacement	A0705	Battery Problem	02/03/2022	02/03/2022
									A070504	Premature Discharge of Battery	02/03/2022	02/03/2022
7672	2022		47265	Dual-chamber implantable pacemaker, rate-responsive	CT141	Electrodes/Leads and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	02/09/2022	02/09/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	02/09/2022	02/09/2022
					CT145	Cardiovascular devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	02/09/2022	02/09/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	02/09/2022	02/09/2022
					CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	02/09/2022	02/09/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	02/09/2022	02/09/2022
					CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	02/09/2022	02/09/2022

							F1901	Additional Surgery	A24	No Apparent Device Problem	02/09/2022	02/09/2022
					CT233	Surgical	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	02/09/2022	02/09/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	02/09/2022	02/09/2022
					CT2406	Implantable	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	02/09/2022	02/09/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	02/09/2022	02/09/2022
					CT2407	Active implantable	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	02/09/2022	02/09/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	02/09/2022	02/09/2022
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	02/09/2022	02/09/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	02/09/2022	02/09/2022
					CT2727	Active implant devices/ancillaries	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	02/09/2022	02/09/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	02/09/2022	02/09/2022
					CT2750	Implantable cardiac electrical stimulators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	02/09/2022	02/09/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	02/09/2022	02/09/2022
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	02/09/2022	02/09/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	02/09/2022	02/09/2022
					CT2985	Active	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	02/09/2022	02/09/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	02/09/2022	02/09/2022
					CT3001	Implantable metal	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	02/09/2022	02/09/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	02/09/2022	02/09/2022
					CT3002	Implantable synthetic polymer	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	02/09/2022	02/09/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	02/09/2022	02/09/2022
					CT3006	Implantable inorganic material	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	02/09/2022	02/09/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	02/09/2022	02/09/2022
					CT314	Electrically powered	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	02/09/2022	02/09/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	02/09/2022	02/09/2022
					CT321	Long-term surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	02/09/2022	02/09/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	02/09/2022	02/09/2022
					CT449	Generators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	02/09/2022	02/09/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	02/09/2022	02/09/2022
					CT981	Single use	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	02/09/2022	02/09/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	02/09/2022	02/09/2022
					CT983	Surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	02/09/2022	02/09/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	02/09/2022	02/09/2022
7673	2022	All other reportable incidents	35852	Single-chamber implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	11/07/2022	11/07/2022
									A070909	Oversensing	11/07/2022	11/07/2022
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	11/07/2022	11/07/2022
									A070909	Oversensing	11/07/2022	11/07/2022
							F1903	Device Explantation	A070504	Premature Discharge of Battery	11/07/2022	11/07/2022
									A070909	Oversensing	11/07/2022	11/07/2022
7674	2022	All other reportable incidents	35852	Single-chamber implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	16/02/2022	16/02/2022
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	16/02/2022	16/02/2022
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	16/02/2022	16/02/2022
7675	2022	All other reportable incidents	35852	Single-chamber implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	31/01/2022	31/01/2022
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	31/01/2022	31/01/2022
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	31/01/2022	31/01/2022
7676	2022	All other reportable incidents	47270	Cardiac resynchronization therapy implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A072201	High Impedance		
							F26	No Health Consequences or Impact	A072201	High Impedance		
7677	2022	All other reportable incidents	36241	Pacemaker/defibrillator lead	CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A072201	High Impedance		
							F26	No Health Consequences or Impact	A072201	High Impedance		
7678	2022	All other reportable incidents	47263	Cardiac resynchronization therapy implantable pacemaker	CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A0908	Incorrect, Inadequate or Imprecise Result or Readings		
									A160503	Inappropriate or Unexpected Reset		
							F26	No Health Consequences or Impact	A0908	Incorrect, Inadequate or Imprecise Result or Readings		
									A160503	Inappropriate or Unexpected Reset		
					CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A0908	Incorrect, Inadequate or Imprecise Result or Readings		
									A160503	Inappropriate or Unexpected Reset		
							F26	No Health Consequences or Impact	A0908	Incorrect, Inadequate or Imprecise Result or Readings		
									A160503	Inappropriate or Unexpected Reset		
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A0908	Incorrect, Inadequate or Imprecise Result or Readings		
									A160503	Inappropriate or Unexpected Reset		
							F26	No Health Consequences or Impact	A0908	Incorrect, Inadequate or Imprecise Result or Readings		
									A160503	Inappropriate or Unexpected Reset		
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A0908	Incorrect, Inadequate or Imprecise Result or Readings		
									A160503	Inappropriate or Unexpected Reset		
							F26	No Health Consequences or Impact	A0908	Incorrect, Inadequate or Imprecise Result or Readings		
									A160503	Inappropriate or Unexpected Reset		
7679	2022	All other reportable incidents	47263	Cardiac resynchronization therapy implantable pacemaker	CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A0908	Incorrect, Inadequate or Imprecise Result or Readings	16/08/2022	16/08/2022
									A160503	Inappropriate or Unexpected Reset	16/08/2022	16/08/2022
							F08	Hospitalization or Prolonged Hospitalization	A0908	Incorrect, Inadequate or Imprecise Result or Readings	16/08/2022	16/08/2022
									A160503	Inappropriate or Unexpected Reset	16/08/2022	16/08/2022
							F1905	Device Revision or Replacement	A0908	Incorrect, Inadequate or Imprecise Result or Readings	16/08/2022	16/08/2022
									A160503	Inappropriate or Unexpected Reset	16/08/2022	16/08/2022
					CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A0908	Incorrect, Inadequate or Imprecise Result or Readings	16/08/2022	16/08/2022
									A160503	Inappropriate or Unexpected Reset	16/08/2022	16/08/2022
							F08	Hospitalization or Prolonged Hospitalization	A0908	Incorrect, Inadequate or Imprecise Result or Readings	16/08/2022	16/08/2022
									A160503	Inappropriate or Unexpected Reset	16/08/2022	16/08/2022
							F1905	Device Revision or Replacement	A0908	Incorrect, Inadequate or Imprecise Result or Readings	16/08/2022	16/08/2022
									A160503	Inappropriate or Unexpected Reset	16/08/2022	16/08/2022

					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A0908	Incorrect, Inadequate or Imprecise Result or Readings	16/08/2022	16/08/2022
									A160503	Inappropriate or Unexpected Reset	16/08/2022	16/08/2022
							F08	Hospitalization or Prolonged Hospitalization	A0908	Incorrect, Inadequate or Imprecise Result or Readings	16/08/2022	16/08/2022
									A160503	Inappropriate or Unexpected Reset	16/08/2022	16/08/2022
							F1905	Device Revision or Replacement	A0908	Incorrect, Inadequate or Imprecise Result or Readings	16/08/2022	16/08/2022
									A160503	Inappropriate or Unexpected Reset	16/08/2022	16/08/2022
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A0908	Incorrect, Inadequate or Imprecise Result or Readings	16/08/2022	16/08/2022
									A160503	Inappropriate or Unexpected Reset	16/08/2022	16/08/2022
							F08	Hospitalization or Prolonged Hospitalization	A0908	Incorrect, Inadequate or Imprecise Result or Readings	16/08/2022	16/08/2022
									A160503	Inappropriate or Unexpected Reset	16/08/2022	16/08/2022
							F1905	Device Revision or Replacement	A0908	Incorrect, Inadequate or Imprecise Result or Readings	16/08/2022	16/08/2022
									A160503	Inappropriate or Unexpected Reset	16/08/2022	16/08/2022
7680	2022	All other reportable incidents	17882	Physiologic-monitoring defibrillation system	CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070803	Failure to Power Up		
							F27	Problem Identified During Non-Clinical Procedure	A070803	Failure to Power Up		
7681	2022	All other reportable incidents	17882	Physiologic-monitoring defibrillation system	CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A1603	Failure of Device to Self-Test		
							F27	Problem Identified During Non-Clinical Procedure	A1603	Failure of Device to Self-Test		
7682	2022	All other reportable incidents	47265	Dual-chamber implantable pacemaker, rate-responsive	CT141	Electrodes/Leads and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
					CT145	Cardiovascular devices	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
					CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
					CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
					CT233	Surgical	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
					CT2406	Implantable	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
					CT2407	Active implantable	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
					CT2727	Active implant devices/ancillaries	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
					CT2750	Implantable cardiac electrical stimulators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
					CT2985	Active	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
					CT3001	Implantable metal	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
					CT3002	Implantable synthetic polymer	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
					CT3006	Implantable inorganic material	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
					CT314	Electrically powered	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
					CT321	Long-term surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A0705	Battery Problem		

									A070504	Premature Discharge of Battery		
				CT449	Generators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem			
								A070504	Premature Discharge of Battery			
						F26	No Health Consequences or Impact	A0705	Battery Problem			
								A070504	Premature Discharge of Battery			
				CT981	Single use	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem			
								A070504	Premature Discharge of Battery			
						F26	No Health Consequences or Impact	A0705	Battery Problem			
								A070504	Premature Discharge of Battery			
				CT983	Surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem			
								A070504	Premature Discharge of Battery			
						F26	No Health Consequences or Impact	A0705	Battery Problem			
								A070504	Premature Discharge of Battery			
7683	2022	Death	47265	Dual-chamber implantable pacemaker, rate-responsive	CT141	Electrodes/Leads and associated devices	E1906	Unspecified Infection	A24	No Apparent Device Problem		
							F02	Death	A24	No Apparent Device Problem		
							F08	Hospitalization or Prolonged Hospitalization	A24	No Apparent Device Problem		
					CT145	Cardiovascular devices	E1906	Unspecified Infection	A24	No Apparent Device Problem		
							F02	Death	A24	No Apparent Device Problem		
							F08	Hospitalization or Prolonged Hospitalization	A24	No Apparent Device Problem		
					CT1684	Cardiac pacemakers and associated devices	E1906	Unspecified Infection	A24	No Apparent Device Problem		
							F02	Death	A24	No Apparent Device Problem		
							F08	Hospitalization or Prolonged Hospitalization	A24	No Apparent Device Problem		
					CT1685	Cardiac pacemakers	E1906	Unspecified Infection	A24	No Apparent Device Problem		
							F02	Death	A24	No Apparent Device Problem		
							F08	Hospitalization or Prolonged Hospitalization	A24	No Apparent Device Problem		
					CT233	Surgical	E1906	Unspecified Infection	A24	No Apparent Device Problem		
							F02	Death	A24	No Apparent Device Problem		
							F08	Hospitalization or Prolonged Hospitalization	A24	No Apparent Device Problem		
					CT2406	Implantable	E1906	Unspecified Infection	A24	No Apparent Device Problem		
							F02	Death	A24	No Apparent Device Problem		
							F08	Hospitalization or Prolonged Hospitalization	A24	No Apparent Device Problem		
					CT2407	Active implantable	E1906	Unspecified Infection	A24	No Apparent Device Problem		
							F02	Death	A24	No Apparent Device Problem		
							F08	Hospitalization or Prolonged Hospitalization	A24	No Apparent Device Problem		
					CT257	Implantable cardiac pacemakers	E1906	Unspecified Infection	A24	No Apparent Device Problem		
							F02	Death	A24	No Apparent Device Problem		
							F08	Hospitalization or Prolonged Hospitalization	A24	No Apparent Device Problem		
					CT2727	Active implant devices/ancillaries	E1906	Unspecified Infection	A24	No Apparent Device Problem		
							F02	Death	A24	No Apparent Device Problem		
							F08	Hospitalization or Prolonged Hospitalization	A24	No Apparent Device Problem		
					CT2750	Implantable cardiac electrical stimulators and associated devices	E1906	Unspecified Infection	A24	No Apparent Device Problem		
							F02	Death	A24	No Apparent Device Problem		
							F08	Hospitalization or Prolonged Hospitalization	A24	No Apparent Device Problem		
					CT2866	Implantable cardiac electrical stimulators	E1906	Unspecified Infection	A24	No Apparent Device Problem		
							F02	Death	A24	No Apparent Device Problem		
							F08	Hospitalization or Prolonged Hospitalization	A24	No Apparent Device Problem		
					CT2985	Active	E1906	Unspecified Infection	A24	No Apparent Device Problem		
							F02	Death	A24	No Apparent Device Problem		
							F08	Hospitalization or Prolonged Hospitalization	A24	No Apparent Device Problem		
					CT3001	Implantable metal	E1906	Unspecified Infection	A24	No Apparent Device Problem		
							F02	Death	A24	No Apparent Device Problem		
							F08	Hospitalization or Prolonged Hospitalization	A24	No Apparent Device Problem		
					CT3002	Implantable synthetic polymer	E1906	Unspecified Infection	A24	No Apparent Device Problem		
							F02	Death	A24	No Apparent Device Problem		
							F08	Hospitalization or Prolonged Hospitalization	A24	No Apparent Device Problem		
					CT3006	Implantable inorganic material	E1906	Unspecified Infection	A24	No Apparent Device Problem		
							F02	Death	A24	No Apparent Device Problem		
							F08	Hospitalization or Prolonged Hospitalization	A24	No Apparent Device Problem		
					CT314	Electrically powered	E1906	Unspecified Infection	A24	No Apparent Device Problem		
							F02	Death	A24	No Apparent Device Problem		
							F08	Hospitalization or Prolonged Hospitalization	A24	No Apparent Device Problem		
					CT321	Long-term surgical invasive	E1906	Unspecified Infection	A24	No Apparent Device Problem		
							F02	Death	A24	No Apparent Device Problem		
							F08	Hospitalization or Prolonged Hospitalization	A24	No Apparent Device Problem		
					CT449	Generators and associated devices	E1906	Unspecified Infection	A24	No Apparent Device Problem		
							F02	Death	A24	No Apparent Device Problem		
							F08	Hospitalization or Prolonged Hospitalization	A24	No Apparent Device Problem		
					CT981	Single use	E1906	Unspecified Infection	A24	No Apparent Device Problem		
							F02	Death	A24	No Apparent Device Problem		
							F08	Hospitalization or Prolonged Hospitalization	A24	No Apparent Device Problem		
					CT983	Surgical invasive	E1906	Unspecified Infection	A24	No Apparent Device Problem		
							F02	Death	A24	No Apparent Device Problem		
							F08	Hospitalization or Prolonged Hospitalization	A24	No Apparent Device Problem		
7684	2022	All other reportable incidents	35852	Single-chamber implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A0508	Noise, Audible		
									A072201	High Impedance		
									A090801	Signal Artifact/Noise		
							F26	No Health Consequences or Impact	A0508	Noise, Audible		
									A072201	High Impedance		
									A090801	Signal Artifact/Noise		
7685	2022	All other reportable incidents	36241	Pacemaker/defibrillator lead	CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A072201	High Impedance		
									A090801	Signal Artifact/Noise		
							F26	No Health Consequences or Impact	A072201	High Impedance		
									A090801	Signal Artifact/Noise		
7686	2022	All other reportable incidents	17882	Physiologic-monitoring defibrillation system	CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A090103	No Audible Prompt/Feedback		
							F26	No Health Consequences or Impact	A090103	No Audible Prompt/Feedback		
7687	2022	All other reportable incidents	47265	Dual-chamber implantable pacemaker, rate-responsive	CT141	Electrodes/Leads and associated devices	E060110	Ventricular Fibrillation	A070504	Premature Discharge of Battery	11/01/2022	11/01/2022
							E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	11/01/2022	11/01/2022
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	11/01/2022	11/01/2022

						F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	11/01/2022	11/01/2022
						F23	Unexpected Medical Intervention	A070504	Premature Discharge of Battery	11/01/2022	11/01/2022
					CT145	Cardiovascular devices	E060110	Ventricular Fibrillation	A070504	Premature Discharge of Battery	11/01/2022	11/01/2022
							E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	11/01/2022	11/01/2022
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	11/01/2022	11/01/2022
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	11/01/2022	11/01/2022
							F23	Unexpected Medical Intervention	A070504	Premature Discharge of Battery	11/01/2022	11/01/2022
					CT1684	Cardiac pacemakers and associated devices	E060110	Ventricular Fibrillation	A070504	Premature Discharge of Battery	11/01/2022	11/01/2022
							E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	11/01/2022	11/01/2022
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	11/01/2022	11/01/2022
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	11/01/2022	11/01/2022
							F23	Unexpected Medical Intervention	A070504	Premature Discharge of Battery	11/01/2022	11/01/2022
					CT1685	Cardiac pacemakers	E060110	Ventricular Fibrillation	A070504	Premature Discharge of Battery	11/01/2022	11/01/2022
							E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	11/01/2022	11/01/2022
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	11/01/2022	11/01/2022
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	11/01/2022	11/01/2022
							F23	Unexpected Medical Intervention	A070504	Premature Discharge of Battery	11/01/2022	11/01/2022
					CT233	Surgical	E060110	Ventricular Fibrillation	A070504	Premature Discharge of Battery	11/01/2022	11/01/2022
							E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	11/01/2022	11/01/2022
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	11/01/2022	11/01/2022
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	11/01/2022	11/01/2022
							F23	Unexpected Medical Intervention	A070504	Premature Discharge of Battery	11/01/2022	11/01/2022
					CT2406	Implantable	E060110	Ventricular Fibrillation	A070504	Premature Discharge of Battery	11/01/2022	11/01/2022
							E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	11/01/2022	11/01/2022
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	11/01/2022	11/01/2022
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	11/01/2022	11/01/2022
							F23	Unexpected Medical Intervention	A070504	Premature Discharge of Battery	11/01/2022	11/01/2022
					CT2407	Active implantable	E060110	Ventricular Fibrillation	A070504	Premature Discharge of Battery	11/01/2022	11/01/2022
							E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	11/01/2022	11/01/2022
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	11/01/2022	11/01/2022
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	11/01/2022	11/01/2022
							F23	Unexpected Medical Intervention	A070504	Premature Discharge of Battery	11/01/2022	11/01/2022
					CT257	Implantable cardiac pacemakers	E060110	Ventricular Fibrillation	A070504	Premature Discharge of Battery	11/01/2022	11/01/2022
							E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	11/01/2022	11/01/2022
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	11/01/2022	11/01/2022
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	11/01/2022	11/01/2022
							F23	Unexpected Medical Intervention	A070504	Premature Discharge of Battery	11/01/2022	11/01/2022
					CT2727	Active implant devices/ancillaries	E060110	Ventricular Fibrillation	A070504	Premature Discharge of Battery	11/01/2022	11/01/2022
							E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	11/01/2022	11/01/2022
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	11/01/2022	11/01/2022
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	11/01/2022	11/01/2022
							F23	Unexpected Medical Intervention	A070504	Premature Discharge of Battery	11/01/2022	11/01/2022
					CT2750	Implantable cardiac electrical stimulators and associated devices	E060110	Ventricular Fibrillation	A070504	Premature Discharge of Battery	11/01/2022	11/01/2022
							E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	11/01/2022	11/01/2022
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	11/01/2022	11/01/2022
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	11/01/2022	11/01/2022
							F23	Unexpected Medical Intervention	A070504	Premature Discharge of Battery	11/01/2022	11/01/2022
					CT2866	Implantable cardiac electrical stimulators	E060110	Ventricular Fibrillation	A070504	Premature Discharge of Battery	11/01/2022	11/01/2022
							E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	11/01/2022	11/01/2022
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	11/01/2022	11/01/2022
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	11/01/2022	11/01/2022
							F23	Unexpected Medical Intervention	A070504	Premature Discharge of Battery	11/01/2022	11/01/2022
					CT2985	Active	E060110	Ventricular Fibrillation	A070504	Premature Discharge of Battery	11/01/2022	11/01/2022
							E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	11/01/2022	11/01/2022
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	11/01/2022	11/01/2022
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	11/01/2022	11/01/2022
							F23	Unexpected Medical Intervention	A070504	Premature Discharge of Battery	11/01/2022	11/01/2022
					CT3001	Implantable metal	E060110	Ventricular Fibrillation	A070504	Premature Discharge of Battery	11/01/2022	11/01/2022
							E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	11/01/2022	11/01/2022
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	11/01/2022	11/01/2022
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	11/01/2022	11/01/2022
							F23	Unexpected Medical Intervention	A070504	Premature Discharge of Battery	11/01/2022	11/01/2022
					CT3002	Implantable synthetic polymer	E060110	Ventricular Fibrillation	A070504	Premature Discharge of Battery	11/01/2022	11/01/2022
							E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	11/01/2022	11/01/2022
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	11/01/2022	11/01/2022
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	11/01/2022	11/01/2022
							F23	Unexpected Medical Intervention	A070504	Premature Discharge of Battery	11/01/2022	11/01/2022
					CT3006	Implantable inorganic material	E060110	Ventricular Fibrillation	A070504	Premature Discharge of Battery	11/01/2022	11/01/2022
							E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	11/01/2022	11/01/2022
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	11/01/2022	11/01/2022
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	11/01/2022	11/01/2022
							F23	Unexpected Medical Intervention	A070504	Premature Discharge of Battery	11/01/2022	11/01/2022
					CT314	Electrically powered	E060110	Ventricular Fibrillation	A070504	Premature Discharge of Battery	11/01/2022	11/01/2022
							E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	11/01/2022	11/01/2022
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	11/01/2022	11/01/2022
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	11/01/2022	11/01/2022
							F23	Unexpected Medical Intervention	A070504	Premature Discharge of Battery	11/01/2022	11/01/2022
					CT321	Long-term surgical invasive	E060110	Ventricular Fibrillation	A070504	Premature Discharge of Battery	11/01/2022	11/01/2022
							E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	11/01/2022	11/01/2022
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	11/01/2022	11/01/2022
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	11/01/2022	11/01/2022
							F23	Unexpected Medical Intervention	A070504	Premature Discharge of Battery	11/01/2022	11/01/2022
					CT449	Generators and associated devices	E060110	Ventricular Fibrillation	A070504	Premature Discharge of Battery	11/01/2022	11/01/2022
							E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	11/01/2022	11/01/2022
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	11/01/2022	11/01/2022
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	11/01/2022	11/01/2022
							F23	Unexpected Medical Intervention	A070504	Premature Discharge of Battery	11/01/2022	11/01/2022
					CT981	Single use	E060110	Ventricular Fibrillation	A070504	Premature Discharge of Battery	11/01/2022	11/01/2022

							E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	11/01/2022	11/01/2022
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	11/01/2022	11/01/2022
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	11/01/2022	11/01/2022
							F23	Unexpected Medical Intervention	A070504	Premature Discharge of Battery	11/01/2022	11/01/2022
					CT983	Surgical invasive	E060110	Ventricular Fibrillation	A070504	Premature Discharge of Battery	11/01/2022	11/01/2022
							E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	11/01/2022	11/01/2022
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	11/01/2022	11/01/2022
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	11/01/2022	11/01/2022
							F23	Unexpected Medical Intervention	A070504	Premature Discharge of Battery	11/01/2022	11/01/2022
7688	2022	All other reportable incidents	47265	Dual-chamber implantable pacemaker, rate-responsive	CT141	Electrodes/Leads and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem	07/02/2022	07/02/2022
							F08	Hospitalization or Prolonged Hospitalization	A0705	Battery Problem	07/02/2022	07/02/2022
							F1905	Device Revision or Replacement	A0705	Battery Problem	07/02/2022	07/02/2022
					CT145	Cardiovascular devices	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem	07/02/2022	07/02/2022
							F08	Hospitalization or Prolonged Hospitalization	A0705	Battery Problem	07/02/2022	07/02/2022
							F1905	Device Revision or Replacement	A0705	Battery Problem	07/02/2022	07/02/2022
					CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem	07/02/2022	07/02/2022
							F08	Hospitalization or Prolonged Hospitalization	A0705	Battery Problem	07/02/2022	07/02/2022
							F1905	Device Revision or Replacement	A0705	Battery Problem	07/02/2022	07/02/2022
					CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem	07/02/2022	07/02/2022
							F08	Hospitalization or Prolonged Hospitalization	A0705	Battery Problem	07/02/2022	07/02/2022
							F1905	Device Revision or Replacement	A0705	Battery Problem	07/02/2022	07/02/2022
					CT233	Surgical	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem	07/02/2022	07/02/2022
							F08	Hospitalization or Prolonged Hospitalization	A0705	Battery Problem	07/02/2022	07/02/2022
							F1905	Device Revision or Replacement	A0705	Battery Problem	07/02/2022	07/02/2022
					CT2406	Implantable	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem	07/02/2022	07/02/2022
							F08	Hospitalization or Prolonged Hospitalization	A0705	Battery Problem	07/02/2022	07/02/2022
							F1905	Device Revision or Replacement	A0705	Battery Problem	07/02/2022	07/02/2022
					CT2407	Active implantable	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem	07/02/2022	07/02/2022
							F08	Hospitalization or Prolonged Hospitalization	A0705	Battery Problem	07/02/2022	07/02/2022
							F1905	Device Revision or Replacement	A0705	Battery Problem	07/02/2022	07/02/2022
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem	07/02/2022	07/02/2022
							F08	Hospitalization or Prolonged Hospitalization	A0705	Battery Problem	07/02/2022	07/02/2022
							F1905	Device Revision or Replacement	A0705	Battery Problem	07/02/2022	07/02/2022
					CT2727	Active implant devices/ancillaries	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem	07/02/2022	07/02/2022
							F08	Hospitalization or Prolonged Hospitalization	A0705	Battery Problem	07/02/2022	07/02/2022
							F1905	Device Revision or Replacement	A0705	Battery Problem	07/02/2022	07/02/2022
					CT2750	Implantable cardiac electrical stimulators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem	07/02/2022	07/02/2022
							F08	Hospitalization or Prolonged Hospitalization	A0705	Battery Problem	07/02/2022	07/02/2022
							F1905	Device Revision or Replacement	A0705	Battery Problem	07/02/2022	07/02/2022
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem	07/02/2022	07/02/2022
							F08	Hospitalization or Prolonged Hospitalization	A0705	Battery Problem	07/02/2022	07/02/2022
							F1905	Device Revision or Replacement	A0705	Battery Problem	07/02/2022	07/02/2022
					CT2985	Active	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem	07/02/2022	07/02/2022
							F08	Hospitalization or Prolonged Hospitalization	A0705	Battery Problem	07/02/2022	07/02/2022
							F1905	Device Revision or Replacement	A0705	Battery Problem	07/02/2022	07/02/2022
					CT3001	Implantable metal	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem	07/02/2022	07/02/2022
							F08	Hospitalization or Prolonged Hospitalization	A0705	Battery Problem	07/02/2022	07/02/2022
							F1905	Device Revision or Replacement	A0705	Battery Problem	07/02/2022	07/02/2022
					CT3002	Implantable synthetic polymer	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem	07/02/2022	07/02/2022
							F08	Hospitalization or Prolonged Hospitalization	A0705	Battery Problem	07/02/2022	07/02/2022
							F1905	Device Revision or Replacement	A0705	Battery Problem	07/02/2022	07/02/2022
					CT3006	Implantable inorganic material	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem	07/02/2022	07/02/2022
							F08	Hospitalization or Prolonged Hospitalization	A0705	Battery Problem	07/02/2022	07/02/2022
							F1905	Device Revision or Replacement	A0705	Battery Problem	07/02/2022	07/02/2022
					CT314	Electrically powered	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem	07/02/2022	07/02/2022
							F08	Hospitalization or Prolonged Hospitalization	A0705	Battery Problem	07/02/2022	07/02/2022
							F1905	Device Revision or Replacement	A0705	Battery Problem	07/02/2022	07/02/2022
					CT321	Long-term surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem	07/02/2022	07/02/2022
							F08	Hospitalization or Prolonged Hospitalization	A0705	Battery Problem	07/02/2022	07/02/2022
							F1905	Device Revision or Replacement	A0705	Battery Problem	07/02/2022	07/02/2022
					CT449	Generators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem	07/02/2022	07/02/2022
							F08	Hospitalization or Prolonged Hospitalization	A0705	Battery Problem	07/02/2022	07/02/2022
							F1905	Device Revision or Replacement	A0705	Battery Problem	07/02/2022	07/02/2022
					CT981	Single use	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem	07/02/2022	07/02/2022
							F08	Hospitalization or Prolonged Hospitalization	A0705	Battery Problem	07/02/2022	07/02/2022
							F1905	Device Revision or Replacement	A0705	Battery Problem	07/02/2022	07/02/2022
					CT983	Surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem	07/02/2022	07/02/2022
							F08	Hospitalization or Prolonged Hospitalization	A0705	Battery Problem	07/02/2022	07/02/2022
							F1905	Device Revision or Replacement	A0705	Battery Problem	07/02/2022	07/02/2022
7689	2022	All other reportable incidents	47265	Dual-chamber implantable pacemaker, rate-responsive	CT141	Electrodes/Leads and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	07/09/2022	07/09/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	07/09/2022	07/09/2022
					CT145	Cardiovascular devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	07/09/2022	07/09/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	07/09/2022	07/09/2022
					CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	07/09/2022	07/09/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	07/09/2022	07/09/2022
					CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	07/09/2022	07/09/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	07/09/2022	07/09/2022
					CT233	Surgical	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	07/09/2022	07/09/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	07/09/2022	07/09/2022
					CT2406	Implantable	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	07/09/2022	07/09/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	07/09/2022	07/09/2022
					CT2407	Active implantable	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	07/09/2022	07/09/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	07/09/2022	07/09/2022
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	07/09/2022	07/09/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	07/09/2022	07/09/2022
					CT2727	Active implant devices/ancillaries	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	07/09/2022	07/09/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	07/09/2022	07/09/2022
					CT2750	Implantable cardiac electrical stimulators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	07/09/2022	07/09/2022

							F1901	Additional Surgery	A24	No Apparent Device Problem	07/09/2022	07/09/2022
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	07/09/2022	07/09/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	07/09/2022	07/09/2022
					CT2985	Active	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	07/09/2022	07/09/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	07/09/2022	07/09/2022
					CT3001	Implantable metal	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	07/09/2022	07/09/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	07/09/2022	07/09/2022
					CT3002	Implantable synthetic polymer	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	07/09/2022	07/09/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	07/09/2022	07/09/2022
					CT3006	Implantable inorganic material	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	07/09/2022	07/09/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	07/09/2022	07/09/2022
					CT314	Electrically powered	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	07/09/2022	07/09/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	07/09/2022	07/09/2022
					CT321	Long-term surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	07/09/2022	07/09/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	07/09/2022	07/09/2022
					CT449	Generators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	07/09/2022	07/09/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	07/09/2022	07/09/2022
					CT981	Single use	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	07/09/2022	07/09/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	07/09/2022	07/09/2022
					CT983	Surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	07/09/2022	07/09/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	07/09/2022	07/09/2022
7690	2022	All other reportable incidents	47265	Dual-chamber implantable pacemaker, rate-responsive	CT141	Electrodes/Leads and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	08/09/2022	08/09/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	08/09/2022	08/09/2022
					CT145	Cardiovascular devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	08/09/2022	08/09/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	08/09/2022	08/09/2022
					CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	08/09/2022	08/09/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	08/09/2022	08/09/2022
					CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	08/09/2022	08/09/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	08/09/2022	08/09/2022
					CT233	Surgical	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	08/09/2022	08/09/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	08/09/2022	08/09/2022
					CT2406	Implantable	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	08/09/2022	08/09/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	08/09/2022	08/09/2022
					CT2407	Active implantable	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	08/09/2022	08/09/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	08/09/2022	08/09/2022
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	08/09/2022	08/09/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	08/09/2022	08/09/2022
					CT2727	Active implant devices/ancillaries	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	08/09/2022	08/09/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	08/09/2022	08/09/2022
					CT2750	Implantable cardiac electrical stimulators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	08/09/2022	08/09/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	08/09/2022	08/09/2022
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	08/09/2022	08/09/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	08/09/2022	08/09/2022
					CT2985	Active	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	08/09/2022	08/09/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	08/09/2022	08/09/2022
					CT3001	Implantable metal	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	08/09/2022	08/09/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	08/09/2022	08/09/2022
					CT3002	Implantable synthetic polymer	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	08/09/2022	08/09/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	08/09/2022	08/09/2022
					CT3006	Implantable inorganic material	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	08/09/2022	08/09/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	08/09/2022	08/09/2022
					CT314	Electrically powered	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	08/09/2022	08/09/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	08/09/2022	08/09/2022
					CT321	Long-term surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	08/09/2022	08/09/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	08/09/2022	08/09/2022
					CT449	Generators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	08/09/2022	08/09/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	08/09/2022	08/09/2022
					CT981	Single use	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	08/09/2022	08/09/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	08/09/2022	08/09/2022
					CT983	Surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	08/09/2022	08/09/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	08/09/2022	08/09/2022
7691	2022		47265	Dual-chamber implantable pacemaker, rate-responsive	CT141	Electrodes/Leads and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	05/09/2022	05/09/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	05/09/2022	05/09/2022
					CT145	Cardiovascular devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	05/09/2022	05/09/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	05/09/2022	05/09/2022
					CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	05/09/2022	05/09/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	05/09/2022	05/09/2022
					CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	05/09/2022	05/09/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	05/09/2022	05/09/2022
					CT233	Surgical	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	05/09/2022	05/09/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	05/09/2022	05/09/2022
					CT2406	Implantable	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	05/09/2022	05/09/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	05/09/2022	05/09/2022
					CT2407	Active implantable	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	05/09/2022	05/09/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	05/09/2022	05/09/2022
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	05/09/2022	05/09/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	05/09/2022	05/09/2022
					CT2727	Active implant devices/ancillaries	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	05/09/2022	05/09/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	05/09/2022	05/09/2022
					CT2750	Implantable cardiac electrical stimulators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	05/09/2022	05/09/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	05/09/2022	05/09/2022
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	05/09/2022	05/09/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	05/09/2022	05/09/2022
					CT2985	Active	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	05/09/2022	05/09/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	05/09/2022	05/09/2022
					CT3001	Implantable metal	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	05/09/2022	05/09/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	05/09/2022	05/09/2022
					CT3002	Implantable synthetic polymer	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	05/09/2022	05/09/2022

						F1901	Additional Surgery	A24	No Apparent Device Problem	05/09/2022	05/09/2022
				CT3006	Implantable inorganic material	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	05/09/2022	05/09/2022
						F1901	Additional Surgery	A24	No Apparent Device Problem	05/09/2022	05/09/2022
				CT314	Electrically powered	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	05/09/2022	05/09/2022
						F1901	Additional Surgery	A24	No Apparent Device Problem	05/09/2022	05/09/2022
				CT321	Long-term surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	05/09/2022	05/09/2022
						F1901	Additional Surgery	A24	No Apparent Device Problem	05/09/2022	05/09/2022
				CT449	Generators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	05/09/2022	05/09/2022
						F1901	Additional Surgery	A24	No Apparent Device Problem	05/09/2022	05/09/2022
				CT981	Single use	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	05/09/2022	05/09/2022
						F1901	Additional Surgery	A24	No Apparent Device Problem	05/09/2022	05/09/2022
				CT983	Surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	05/09/2022	05/09/2022
						F1901	Additional Surgery	A24	No Apparent Device Problem	05/09/2022	05/09/2022
7692	2022	All other reportable incidents	47265	Dual-chamber implantable pacemaker, rate-responsive	CT141	Electrodes/Leads and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	22/08/2022	22/08/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	22/08/2022	22/08/2022
					CT145	Cardiovascular devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	22/08/2022	22/08/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	22/08/2022	22/08/2022
					CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	22/08/2022	22/08/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	22/08/2022	22/08/2022
					CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	22/08/2022	22/08/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	22/08/2022	22/08/2022
					CT233	Surgical	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	22/08/2022	22/08/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	22/08/2022	22/08/2022
					CT2406	Implantable	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	22/08/2022	22/08/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	22/08/2022	22/08/2022
					CT2407	Active implantable	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	22/08/2022	22/08/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	22/08/2022	22/08/2022
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	22/08/2022	22/08/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	22/08/2022	22/08/2022
					CT2727	Active implant devices/ancillaries	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	22/08/2022	22/08/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	22/08/2022	22/08/2022
					CT2750	Implantable cardiac electrical stimulators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	22/08/2022	22/08/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	22/08/2022	22/08/2022
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	22/08/2022	22/08/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	22/08/2022	22/08/2022
					CT2985	Active	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	22/08/2022	22/08/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	22/08/2022	22/08/2022
					CT3001	Implantable metal	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	22/08/2022	22/08/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	22/08/2022	22/08/2022
					CT3002	Implantable synthetic polymer	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	22/08/2022	22/08/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	22/08/2022	22/08/2022
					CT3006	Implantable inorganic material	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	22/08/2022	22/08/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	22/08/2022	22/08/2022
					CT314	Electrically powered	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	22/08/2022	22/08/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	22/08/2022	22/08/2022
					CT321	Long-term surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	22/08/2022	22/08/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	22/08/2022	22/08/2022
					CT449	Generators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	22/08/2022	22/08/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	22/08/2022	22/08/2022
					CT981	Single use	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	22/08/2022	22/08/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	22/08/2022	22/08/2022
					CT983	Surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	22/08/2022	22/08/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	22/08/2022	22/08/2022
7693	2022	All other reportable incidents	37265	Dual-chamber implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E2104	Electric Shock	A071302	Inappropriate/Inadequate Shock/Stimulation	19/01/2022	19/01/2022
							E2403	No Clinical Signs, Symptoms or Conditions	A071302	Inappropriate/Inadequate Shock/Stimulation	19/01/2022	19/01/2022
							F08	Hospitalization or Prolonged Hospitalization	A071302	Inappropriate/Inadequate Shock/Stimulation	19/01/2022	19/01/2022
							F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A071302	Inappropriate/Inadequate Shock/Stimulation	19/01/2022	19/01/2022
7694	2022	All other reportable incidents	47267	Single-chamber implantable pacemaker, rate-responsive	CT141	Electrodes/Leads and associated devices	E060109	Tachycardia	A070909	Oversensing		
									A0712	Pacing Problem		
									A072001	Electromagnetic Interference		
									A090801	Signal Artifact/Noise		
							E2403	No Clinical Signs, Symptoms or Conditions	A070909	Oversensing		
									A0712	Pacing Problem		
									A072001	Electromagnetic Interference		
									A090801	Signal Artifact/Noise		
							F26	No Health Consequences or Impact	A070909	Oversensing		
									A0712	Pacing Problem		
									A072001	Electromagnetic Interference		
									A090801	Signal Artifact/Noise		
					CT145	Cardiovascular devices	E060109	Tachycardia	A070909	Oversensing		
									A0712	Pacing Problem		
									A072001	Electromagnetic Interference		
									A090801	Signal Artifact/Noise		
							E2403	No Clinical Signs, Symptoms or Conditions	A070909	Oversensing		
									A0712	Pacing Problem		
									A072001	Electromagnetic Interference		
									A090801	Signal Artifact/Noise		
							F26	No Health Consequences or Impact	A070909	Oversensing		
									A0712	Pacing Problem		
									A072001	Electromagnetic Interference		
									A090801	Signal Artifact/Noise		
					CT1684	Cardiac pacemakers and associated devices	E060109	Tachycardia	A070909	Oversensing		
									A0712	Pacing Problem		
									A072001	Electromagnetic Interference		
									A090801	Signal Artifact/Noise		
							E2403	No Clinical Signs, Symptoms or Conditions	A070909	Oversensing		
									A0712	Pacing Problem		
									A072001	Electromagnetic Interference		

									A090801	Signal Artifact/Noise			
							F26	No Health Consequences or Impact	A070909	Oversensing			
									A0712	Pacing Problem			
									A072001	Electromagnetic Interference			
									A090801	Signal Artifact/Noise			
					CT1685	Cardiac pacemakers	E060109	Tachycardia	A070909	Oversensing			
									A0712	Pacing Problem			
									A072001	Electromagnetic Interference			
									A090801	Signal Artifact/Noise			
							E2403	No Clinical Signs, Symptoms or Conditions	A070909	Oversensing			
									A0712	Pacing Problem			
									A072001	Electromagnetic Interference			
									A090801	Signal Artifact/Noise			
							F26	No Health Consequences or Impact	A070909	Oversensing			
									A0712	Pacing Problem			
									A072001	Electromagnetic Interference			
									A090801	Signal Artifact/Noise			
					CT233	Surgical	E060109	Tachycardia	A070909	Oversensing			
									A0712	Pacing Problem			
									A072001	Electromagnetic Interference			
									A090801	Signal Artifact/Noise			
							E2403	No Clinical Signs, Symptoms or Conditions	A070909	Oversensing			
									A0712	Pacing Problem			
									A072001	Electromagnetic Interference			
									A090801	Signal Artifact/Noise			
							F26	No Health Consequences or Impact	A070909	Oversensing			
									A0712	Pacing Problem			
									A072001	Electromagnetic Interference			
									A090801	Signal Artifact/Noise			
					CT2406	Implantable	E060109	Tachycardia	A070909	Oversensing			
									A0712	Pacing Problem			
									A072001	Electromagnetic Interference			
									A090801	Signal Artifact/Noise			
							E2403	No Clinical Signs, Symptoms or Conditions	A070909	Oversensing			
									A0712	Pacing Problem			
									A072001	Electromagnetic Interference			
									A090801	Signal Artifact/Noise			
							F26	No Health Consequences or Impact	A070909	Oversensing			
									A0712	Pacing Problem			
									A072001	Electromagnetic Interference			
									A090801	Signal Artifact/Noise			
					CT2407	Active implantable	E060109	Tachycardia	A070909	Oversensing			
									A0712	Pacing Problem			
									A072001	Electromagnetic Interference			
									A090801	Signal Artifact/Noise			
							E2403	No Clinical Signs, Symptoms or Conditions	A070909	Oversensing			
									A0712	Pacing Problem			
									A072001	Electromagnetic Interference			
									A090801	Signal Artifact/Noise			
							F26	No Health Consequences or Impact	A070909	Oversensing			
									A0712	Pacing Problem			
									A072001	Electromagnetic Interference			
									A090801	Signal Artifact/Noise			
					CT257	Implantable cardiac pacemakers	E060109	Tachycardia	A070909	Oversensing			
									A0712	Pacing Problem			
									A072001	Electromagnetic Interference			
									A090801	Signal Artifact/Noise			
							E2403	No Clinical Signs, Symptoms or Conditions	A070909	Oversensing			
									A0712	Pacing Problem			
									A072001	Electromagnetic Interference			
									A090801	Signal Artifact/Noise			
							F26	No Health Consequences or Impact	A070909	Oversensing			
									A0712	Pacing Problem			
									A072001	Electromagnetic Interference			
									A090801	Signal Artifact/Noise			
					CT2727	Active implant devices/ancillaries	E060109	Tachycardia	A070909	Oversensing			
									A0712	Pacing Problem			
									A072001	Electromagnetic Interference			
									A090801	Signal Artifact/Noise			
							E2403	No Clinical Signs, Symptoms or Conditions	A070909	Oversensing			
									A0712	Pacing Problem			
									A072001	Electromagnetic Interference			
									A090801	Signal Artifact/Noise			
							F26	No Health Consequences or Impact	A070909	Oversensing			
									A0712	Pacing Problem			
									A072001	Electromagnetic Interference			
									A090801	Signal Artifact/Noise			
					CT2750	Implantable cardiac electrical stimulators and associated devices	E060109	Tachycardia	A070909	Oversensing			
									A0712	Pacing Problem			
									A072001	Electromagnetic Interference			
									A090801	Signal Artifact/Noise			
							E2403	No Clinical Signs, Symptoms or Conditions	A070909	Oversensing			
									A0712	Pacing Problem			
									A072001	Electromagnetic Interference			
									A090801	Signal Artifact/Noise			
							F26	No Health Consequences or Impact	A070909	Oversensing			
									A0712	Pacing Problem			
									A072001	Electromagnetic Interference			

									A090801	Signal Artifact/Noise		
					CT2866	Implantable cardiac electrical stimulators	E060109	Tachycardia	A070909	Oversensing		
									A0712	Pacing Problem		
									A072001	Electromagnetic Interference		
									A090801	Signal Artifact/Noise		
							E2403	No Clinical Signs, Symptoms or Conditions	A070909	Oversensing		
									A0712	Pacing Problem		
									A072001	Electromagnetic Interference		
									A090801	Signal Artifact/Noise		
							F26	No Health Consequences or Impact	A070909	Oversensing		
									A0712	Pacing Problem		
									A072001	Electromagnetic Interference		
									A090801	Signal Artifact/Noise		
					CT2985	Active	E060109	Tachycardia	A070909	Oversensing		
									A0712	Pacing Problem		
									A072001	Electromagnetic Interference		
									A090801	Signal Artifact/Noise		
							E2403	No Clinical Signs, Symptoms or Conditions	A070909	Oversensing		
									A0712	Pacing Problem		
									A072001	Electromagnetic Interference		
									A090801	Signal Artifact/Noise		
							F26	No Health Consequences or Impact	A070909	Oversensing		
									A0712	Pacing Problem		
									A072001	Electromagnetic Interference		
									A090801	Signal Artifact/Noise		
					CT3001	Implantable metal	E060109	Tachycardia	A070909	Oversensing		
									A0712	Pacing Problem		
									A072001	Electromagnetic Interference		
									A090801	Signal Artifact/Noise		
							E2403	No Clinical Signs, Symptoms or Conditions	A070909	Oversensing		
									A0712	Pacing Problem		
									A072001	Electromagnetic Interference		
									A090801	Signal Artifact/Noise		
							F26	No Health Consequences or Impact	A070909	Oversensing		
									A0712	Pacing Problem		
									A072001	Electromagnetic Interference		
									A090801	Signal Artifact/Noise		
					CT3002	Implantable synthetic polymer	E060109	Tachycardia	A070909	Oversensing		
									A0712	Pacing Problem		
									A072001	Electromagnetic Interference		
									A090801	Signal Artifact/Noise		
							E2403	No Clinical Signs, Symptoms or Conditions	A070909	Oversensing		
									A0712	Pacing Problem		
									A072001	Electromagnetic Interference		
									A090801	Signal Artifact/Noise		
							F26	No Health Consequences or Impact	A070909	Oversensing		
									A0712	Pacing Problem		
									A072001	Electromagnetic Interference		
									A090801	Signal Artifact/Noise		
					CT3006	Implantable inorganic material	E060109	Tachycardia	A070909	Oversensing		
									A0712	Pacing Problem		
									A072001	Electromagnetic Interference		
									A090801	Signal Artifact/Noise		
							E2403	No Clinical Signs, Symptoms or Conditions	A070909	Oversensing		
									A0712	Pacing Problem		
									A072001	Electromagnetic Interference		
									A090801	Signal Artifact/Noise		
							F26	No Health Consequences or Impact	A070909	Oversensing		
									A0712	Pacing Problem		
									A072001	Electromagnetic Interference		
									A090801	Signal Artifact/Noise		
					CT314	Electrically powered	E060109	Tachycardia	A070909	Oversensing		
									A0712	Pacing Problem		
									A072001	Electromagnetic Interference		
									A090801	Signal Artifact/Noise		
							E2403	No Clinical Signs, Symptoms or Conditions	A070909	Oversensing		
									A0712	Pacing Problem		
									A072001	Electromagnetic Interference		
									A090801	Signal Artifact/Noise		
							F26	No Health Consequences or Impact	A070909	Oversensing		
									A0712	Pacing Problem		
									A072001	Electromagnetic Interference		
									A090801	Signal Artifact/Noise		
					CT321	Long-term surgical invasive	E060109	Tachycardia	A070909	Oversensing		
									A0712	Pacing Problem		
									A072001	Electromagnetic Interference		
									A090801	Signal Artifact/Noise		
							E2403	No Clinical Signs, Symptoms or Conditions	A070909	Oversensing		
									A0712	Pacing Problem		
									A072001	Electromagnetic Interference		
									A090801	Signal Artifact/Noise		
							F26	No Health Consequences or Impact	A070909	Oversensing		
									A0712	Pacing Problem		
									A072001	Electromagnetic Interference		
									A090801	Signal Artifact/Noise		
					CT449	Generators and associated devices	E060109	Tachycardia	A070909	Oversensing		
									A0712	Pacing Problem		
									A072001	Electromagnetic Interference		

									A090801	Signal Artifact/Noise		
						E2403	No Clinical Signs, Symptoms or Conditions	A070909	Oversensing		
								A0712	Pacing Problem		
								A072001	Electromagnetic Interference		
								A090801	Signal Artifact/Noise		
						F26	No Health Consequences or Impact	A070909	Oversensing		
								A0712	Pacing Problem		
								A072001	Electromagnetic Interference		
								A090801	Signal Artifact/Noise		
					CT981	Single use	E060109	Tachycardia	A070909	Oversensing		
									A0712	Pacing Problem		
									A072001	Electromagnetic Interference		
									A090801	Signal Artifact/Noise		
						E2403	No Clinical Signs, Symptoms or Conditions	A070909	Oversensing		
								A0712	Pacing Problem		
								A072001	Electromagnetic Interference		
								A090801	Signal Artifact/Noise		
						F26	No Health Consequences or Impact	A070909	Oversensing		
								A0712	Pacing Problem		
								A072001	Electromagnetic Interference		
								A090801	Signal Artifact/Noise		
					CT983	Surgical invasive	E060109	Tachycardia	A070909	Oversensing		
									A0712	Pacing Problem		
									A072001	Electromagnetic Interference		
									A090801	Signal Artifact/Noise		
						E2403	No Clinical Signs, Symptoms or Conditions	A070909	Oversensing		
								A0712	Pacing Problem		
								A072001	Electromagnetic Interference		
								A090801	Signal Artifact/Noise		
						F26	No Health Consequences or Impact	A070909	Oversensing		
								A0712	Pacing Problem		
								A072001	Electromagnetic Interference		
								A090801	Signal Artifact/Noise		
7695	2022	All other reportable incidents	47263	Cardiac resynchronization therapy implantable pacemaker	CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A0908	Incorrect, Inadequate or Imprecise Result or Readings	29/12/2021	29/12/2021
							F08	Hospitalization or Prolonged Hospitalization	A0908	Incorrect, Inadequate or Imprecise Result or Readings	29/12/2021	29/12/2021
							F1905	Device Revision or Replacement	A0908	Incorrect, Inadequate or Imprecise Result or Readings	29/12/2021	29/12/2021
					CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A0908	Incorrect, Inadequate or Imprecise Result or Readings	29/12/2021	29/12/2021
							F08	Hospitalization or Prolonged Hospitalization	A0908	Incorrect, Inadequate or Imprecise Result or Readings	29/12/2021	29/12/2021
							F1905	Device Revision or Replacement	A0908	Incorrect, Inadequate or Imprecise Result or Readings	29/12/2021	29/12/2021
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A0908	Incorrect, Inadequate or Imprecise Result or Readings	29/12/2021	29/12/2021
							F08	Hospitalization or Prolonged Hospitalization	A0908	Incorrect, Inadequate or Imprecise Result or Readings	29/12/2021	29/12/2021
							F1905	Device Revision or Replacement	A0908	Incorrect, Inadequate or Imprecise Result or Readings	29/12/2021	29/12/2021
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A0908	Incorrect, Inadequate or Imprecise Result or Readings	29/12/2021	29/12/2021
							F08	Hospitalization or Prolonged Hospitalization	A0908	Incorrect, Inadequate or Imprecise Result or Readings	29/12/2021	29/12/2021
							F1905	Device Revision or Replacement	A0908	Incorrect, Inadequate or Imprecise Result or Readings	29/12/2021	29/12/2021
7696	2022	All other reportable incidents	35223	Endocardial/interventricular septal pacing lead	CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070909	Oversensing		
							F1905	Device Revision or Replacement	A070909	Oversensing		
7697	2022	All other reportable incidents	47263	Cardiac resynchronization therapy implantable pacemaker	CT1684	Cardiac pacemakers and associated devices	E2103	Device Overstimulation of Tissue	A0908	Incorrect, Inadequate or Imprecise Result or Readings	16/08/2022	16/08/2022
									A0909	Unexpected Therapeutic Results	16/08/2022	16/08/2022
							E2403	No Clinical Signs, Symptoms or Conditions	A0908	Incorrect, Inadequate or Imprecise Result or Readings	16/08/2022	16/08/2022
									A0909	Unexpected Therapeutic Results	16/08/2022	16/08/2022
							F08	Hospitalization or Prolonged Hospitalization	A0908	Incorrect, Inadequate or Imprecise Result or Readings	16/08/2022	16/08/2022
									A0909	Unexpected Therapeutic Results	16/08/2022	16/08/2022
							F1905	Device Revision or Replacement	A0908	Incorrect, Inadequate or Imprecise Result or Readings	16/08/2022	16/08/2022
									A0909	Unexpected Therapeutic Results	16/08/2022	16/08/2022
					CT1685	Cardiac pacemakers	E2103	Device Overstimulation of Tissue	A0908	Incorrect, Inadequate or Imprecise Result or Readings	16/08/2022	16/08/2022
									A0909	Unexpected Therapeutic Results	16/08/2022	16/08/2022
							E2403	No Clinical Signs, Symptoms or Conditions	A0908	Incorrect, Inadequate or Imprecise Result or Readings	16/08/2022	16/08/2022
									A0909	Unexpected Therapeutic Results	16/08/2022	16/08/2022
							F08	Hospitalization or Prolonged Hospitalization	A0908	Incorrect, Inadequate or Imprecise Result or Readings	16/08/2022	16/08/2022
									A0909	Unexpected Therapeutic Results	16/08/2022	16/08/2022
							F1905	Device Revision or Replacement	A0908	Incorrect, Inadequate or Imprecise Result or Readings	16/08/2022	16/08/2022
									A0909	Unexpected Therapeutic Results	16/08/2022	16/08/2022
					CT257	Implantable cardiac pacemakers	E2103	Device Overstimulation of Tissue	A0908	Incorrect, Inadequate or Imprecise Result or Readings	16/08/2022	16/08/2022
									A0909	Unexpected Therapeutic Results	16/08/2022	16/08/2022
							E2403	No Clinical Signs, Symptoms or Conditions	A0908	Incorrect, Inadequate or Imprecise Result or Readings	16/08/2022	16/08/2022
									A0909	Unexpected Therapeutic Results	16/08/2022	16/08/2022
							F08	Hospitalization or Prolonged Hospitalization	A0908	Incorrect, Inadequate or Imprecise Result or Readings	16/08/2022	16/08/2022
									A0909	Unexpected Therapeutic Results	16/08/2022	16/08/2022

							F1905	Device Revision or Replacement	A0908	Incorrect, Inadequate or Imprecise Result or Readings	16/08/2022	16/08/2022
									A0909	Unexpected Therapeutic Results	16/08/2022	16/08/2022
					CT2866	Implantable cardiac electrical stimulators	E2103	Device Overstimulation of Tissue	A0908	Incorrect, Inadequate or Imprecise Result or Readings	16/08/2022	16/08/2022
									A0909	Unexpected Therapeutic Results	16/08/2022	16/08/2022
							E2403	No Clinical Signs, Symptoms or Conditions	A0908	Incorrect, Inadequate or Imprecise Result or Readings	16/08/2022	16/08/2022
									A0909	Unexpected Therapeutic Results	16/08/2022	16/08/2022
							F08	Hospitalization or Prolonged Hospitalization	A0908	Incorrect, Inadequate or Imprecise Result or Readings	16/08/2022	16/08/2022
									A0909	Unexpected Therapeutic Results	16/08/2022	16/08/2022
							F1905	Device Revision or Replacement	A0908	Incorrect, Inadequate or Imprecise Result or Readings	16/08/2022	16/08/2022
									A0909	Unexpected Therapeutic Results	16/08/2022	16/08/2022
7698	2022	All other reportable incidents	17882	Physiologic-monitoring defibrillation system	CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070803	Failure to Power Up		
									A0719	Unexpected Shutdown		
							F26	No Health Consequences or Impact	A070803	Failure to Power Up		
									A0719	Unexpected Shutdown		
7699	2022	All other reportable incidents	47263	Cardiac resynchronization therapy implantable pacemaker	CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	14/06/2022	14/06/2022
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	14/06/2022	14/06/2022
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	14/06/2022	14/06/2022
					CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	14/06/2022	14/06/2022
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	14/06/2022	14/06/2022
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	14/06/2022	14/06/2022
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	14/06/2022	14/06/2022
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	14/06/2022	14/06/2022
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	14/06/2022	14/06/2022
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	14/06/2022	14/06/2022
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	14/06/2022	14/06/2022
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	14/06/2022	14/06/2022
7700	2022		47265	Dual-chamber implantable pacemaker, rate-responsive	CT141	Electrodes/Leads and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	10/09/2022	10/09/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	10/09/2022	10/09/2022
					CT145	Cardiovascular devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	10/09/2022	10/09/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	10/09/2022	10/09/2022
					CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	10/09/2022	10/09/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	10/09/2022	10/09/2022
					CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	10/09/2022	10/09/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	10/09/2022	10/09/2022
					CT233	Surgical	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	10/09/2022	10/09/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	10/09/2022	10/09/2022
					CT2406	Implantable	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	10/09/2022	10/09/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	10/09/2022	10/09/2022
					CT2407	Active implantable	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	10/09/2022	10/09/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	10/09/2022	10/09/2022
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	10/09/2022	10/09/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	10/09/2022	10/09/2022
					CT2727	Active implant devices/ancillaries	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	10/09/2022	10/09/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	10/09/2022	10/09/2022
					CT2750	Implantable cardiac electrical stimulators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	10/09/2022	10/09/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	10/09/2022	10/09/2022
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	10/09/2022	10/09/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	10/09/2022	10/09/2022
					CT2985	Active	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	10/09/2022	10/09/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	10/09/2022	10/09/2022
					CT3001	Implantable metal	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	10/09/2022	10/09/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	10/09/2022	10/09/2022
					CT3002	Implantable synthetic polymer	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	10/09/2022	10/09/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	10/09/2022	10/09/2022
					CT3006	Implantable inorganic material	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	10/09/2022	10/09/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	10/09/2022	10/09/2022
					CT314	Electrically powered	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	10/09/2022	10/09/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	10/09/2022	10/09/2022
					CT321	Long-term surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	10/09/2022	10/09/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	10/09/2022	10/09/2022
					CT449	Generators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	10/09/2022	10/09/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	10/09/2022	10/09/2022
					CT981	Single use	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	10/09/2022	10/09/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	10/09/2022	10/09/2022
					CT983	Surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	10/09/2022	10/09/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	10/09/2022	10/09/2022
7701	2022	All other reportable incidents	47265	Dual-chamber implantable pacemaker, rate-responsive	CT141	Electrodes/Leads and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	08/09/2022	08/09/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	08/09/2022	08/09/2022
					CT145	Cardiovascular devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	08/09/2022	08/09/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	08/09/2022	08/09/2022
					CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	08/09/2022	08/09/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	08/09/2022	08/09/2022
					CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	08/09/2022	08/09/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	08/09/2022	08/09/2022
					CT233	Surgical	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	08/09/2022	08/09/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	08/09/2022	08/09/2022
					CT2406	Implantable	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	08/09/2022	08/09/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	08/09/2022	08/09/2022
					CT2407	Active implantable	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	08/09/2022	08/09/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	08/09/2022	08/09/2022
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	08/09/2022	08/09/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	08/09/2022	08/09/2022
					CT2727	Active implant devices/ancillaries	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	08/09/2022	08/09/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	08/09/2022	08/09/2022

					CT2750	Implantable cardiac electrical stimulators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	08/09/2022	08/09/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	08/09/2022	08/09/2022
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	08/09/2022	08/09/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	08/09/2022	08/09/2022
					CT2985	Active	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	08/09/2022	08/09/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	08/09/2022	08/09/2022
					CT3001	Implantable metal	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	08/09/2022	08/09/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	08/09/2022	08/09/2022
					CT3002	Implantable synthetic polymer	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	08/09/2022	08/09/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	08/09/2022	08/09/2022
					CT3006	Implantable inorganic material	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	08/09/2022	08/09/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	08/09/2022	08/09/2022
					CT314	Electrically powered	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	08/09/2022	08/09/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	08/09/2022	08/09/2022
					CT321	Long-term surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	08/09/2022	08/09/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	08/09/2022	08/09/2022
					CT449	Generators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	08/09/2022	08/09/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	08/09/2022	08/09/2022
					CT981	Single use	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	08/09/2022	08/09/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	08/09/2022	08/09/2022
					CT983	Surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	08/09/2022	08/09/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	08/09/2022	08/09/2022
7702	2022	All other reportable incidents	47265	Dual-chamber implantable pacemaker, rate-responsive	CT141	Electrodes/Leads and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	08/09/2022	08/09/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	08/09/2022	08/09/2022
					CT145	Cardiovascular devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	08/09/2022	08/09/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	08/09/2022	08/09/2022
					CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	08/09/2022	08/09/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	08/09/2022	08/09/2022
					CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	08/09/2022	08/09/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	08/09/2022	08/09/2022
					CT233	Surgical	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	08/09/2022	08/09/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	08/09/2022	08/09/2022
					CT2406	Implantable	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	08/09/2022	08/09/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	08/09/2022	08/09/2022
					CT2407	Active implantable	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	08/09/2022	08/09/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	08/09/2022	08/09/2022
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	08/09/2022	08/09/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	08/09/2022	08/09/2022
					CT2727	Active implant devices/ancillaries	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	08/09/2022	08/09/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	08/09/2022	08/09/2022
					CT2750	Implantable cardiac electrical stimulators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	08/09/2022	08/09/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	08/09/2022	08/09/2022
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	08/09/2022	08/09/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	08/09/2022	08/09/2022
					CT2985	Active	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	08/09/2022	08/09/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	08/09/2022	08/09/2022
					CT3001	Implantable metal	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	08/09/2022	08/09/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	08/09/2022	08/09/2022
					CT3002	Implantable synthetic polymer	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	08/09/2022	08/09/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	08/09/2022	08/09/2022
					CT3006	Implantable inorganic material	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	08/09/2022	08/09/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	08/09/2022	08/09/2022
					CT314	Electrically powered	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	08/09/2022	08/09/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	08/09/2022	08/09/2022
					CT321	Long-term surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	08/09/2022	08/09/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	08/09/2022	08/09/2022
					CT449	Generators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	08/09/2022	08/09/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	08/09/2022	08/09/2022
					CT981	Single use	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	08/09/2022	08/09/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	08/09/2022	08/09/2022
					CT983	Surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	08/09/2022	08/09/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	08/09/2022	08/09/2022
7703	2022	All other reportable incidents	47265	Dual-chamber implantable pacemaker, rate-responsive	CT141	Electrodes/Leads and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	08/09/2022	08/09/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	08/09/2022	08/09/2022
					CT145	Cardiovascular devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	08/09/2022	08/09/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	08/09/2022	08/09/2022
					CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	08/09/2022	08/09/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	08/09/2022	08/09/2022
					CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	08/09/2022	08/09/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	08/09/2022	08/09/2022
					CT233	Surgical	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	08/09/2022	08/09/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	08/09/2022	08/09/2022
					CT2406	Implantable	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	08/09/2022	08/09/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	08/09/2022	08/09/2022
					CT2407	Active implantable	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	08/09/2022	08/09/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	08/09/2022	08/09/2022
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	08/09/2022	08/09/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	08/09/2022	08/09/2022
					CT2727	Active implant devices/ancillaries	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	08/09/2022	08/09/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	08/09/2022	08/09/2022
					CT2750	Implantable cardiac electrical stimulators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	08/09/2022	08/09/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	08/09/2022	08/09/2022
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	08/09/2022	08/09/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	08/09/2022	08/09/2022
					CT2985	Active	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	08/09/2022	08/09/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	08/09/2022	08/09/2022
					CT3001	Implantable metal	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	08/09/2022	08/09/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	08/09/2022	08/09/2022

					CT3002	Implantable synthetic polymer	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	08/09/2022	08/09/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	08/09/2022	08/09/2022
					CT3006	Implantable inorganic material	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	08/09/2022	08/09/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	08/09/2022	08/09/2022
					CT314	Electrically powered	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	08/09/2022	08/09/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	08/09/2022	08/09/2022
					CT321	Long-term surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	08/09/2022	08/09/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	08/09/2022	08/09/2022
					CT449	Generators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	08/09/2022	08/09/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	08/09/2022	08/09/2022
					CT981	Single use	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	08/09/2022	08/09/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	08/09/2022	08/09/2022
					CT983	Surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	08/09/2022	08/09/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	08/09/2022	08/09/2022
7704	2022	All other reportable incidents	47263	Cardiac resynchronization therapy implantable pacemaker	CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A160601	Premature Elective Replacement Indicator	05/09/2022	05/09/2022
							F1903	Device Explantation	A160601	Premature Elective Replacement Indicator	05/09/2022	05/09/2022
					CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A160601	Premature Elective Replacement Indicator	05/09/2022	05/09/2022
							F1903	Device Explantation	A160601	Premature Elective Replacement Indicator	05/09/2022	05/09/2022
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A160601	Premature Elective Replacement Indicator	05/09/2022	05/09/2022
							F1903	Device Explantation	A160601	Premature Elective Replacement Indicator	05/09/2022	05/09/2022
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A160601	Premature Elective Replacement Indicator	05/09/2022	05/09/2022
							F1903	Device Explantation	A160601	Premature Elective Replacement Indicator	05/09/2022	05/09/2022
7705	2022	All other reportable incidents	35223	Endocardial/interventricular septal pacing lead	CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070909	Oversensing		
									A070910	Undersensing		
									A090801	Signal Artifact/Noise		
									A12	Connection Problem		
							F26	No Health Consequences or Impact	A070909	Oversensing		
									A070910	Undersensing		
									A090801	Signal Artifact/Noise		
									A12	Connection Problem		
7706	2022	All other reportable incidents	17882	Physiologic-monitoring defibrillation system	CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A07	Electrical /Electronic Property Problem		
							F26	No Health Consequences or Impact	A07	Electrical /Electronic Property Problem		
7707	2022	All other reportable incidents	35852	Single-chamber implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	26/01/2022	26/01/2022
							F1901	Additional Surgery	A070504	Premature Discharge of Battery	26/01/2022	26/01/2022
7708	2022	All other reportable incidents	35852	Single-chamber implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E2104	Electric Shock	A072201	High Impedance		
							E2403	No Clinical Signs, Symptoms or Conditions	A072201	High Impedance		
							F17	Sedation	A072201	High Impedance		
							F2203	Imaging Required	A072201	High Impedance		
							F26	No Health Consequences or Impact	A072201	High Impedance		
7709	2022	All other reportable incidents	17882	Physiologic-monitoring defibrillation system	CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A090812	Missing Test Results		
							F27	Problem Identified During Non-Clinical Procedure	A090812	Missing Test Results		
7710	2022	All other reportable incidents	47265	Dual-chamber implantable pacemaker, rate-responsive	CT141	Electrodes/Leads and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
					CT145	Cardiovascular devices	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
					CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
					CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
					CT233	Surgical	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
					CT2406	Implantable	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
					CT2407	Active implantable	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
					CT2727	Active implant devices/ancillaries	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
					CT2750	Implantable cardiac electrical stimulators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
					CT2985	Active	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A0705	Battery Problem		

									A070504	Premature Discharge of Battery			
					CT3001	Implantable metal	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem			
									A070504	Premature Discharge of Battery			
							F26	No Health Consequences or Impact	A0705	Battery Problem			
									A070504	Premature Discharge of Battery			
					CT3002	Implantable synthetic polymer	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem			
									A070504	Premature Discharge of Battery			
							F26	No Health Consequences or Impact	A0705	Battery Problem			
									A070504	Premature Discharge of Battery			
					CT3006	Implantable inorganic material	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem			
									A070504	Premature Discharge of Battery			
							F26	No Health Consequences or Impact	A0705	Battery Problem			
									A070504	Premature Discharge of Battery			
					CT314	Electrically powered	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem			
									A070504	Premature Discharge of Battery			
							F26	No Health Consequences or Impact	A0705	Battery Problem			
									A070504	Premature Discharge of Battery			
					CT321	Long-term surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem			
									A070504	Premature Discharge of Battery			
							F26	No Health Consequences or Impact	A0705	Battery Problem			
									A070504	Premature Discharge of Battery			
					CT449	Generators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem			
									A070504	Premature Discharge of Battery			
							F26	No Health Consequences or Impact	A0705	Battery Problem			
									A070504	Premature Discharge of Battery			
					CT981	Single use	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem			
									A070504	Premature Discharge of Battery			
							F26	No Health Consequences or Impact	A0705	Battery Problem			
									A070504	Premature Discharge of Battery			
					CT983	Surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem			
									A070504	Premature Discharge of Battery			
							F26	No Health Consequences or Impact	A0705	Battery Problem			
									A070504	Premature Discharge of Battery			
7711	2022	All other reportable incidents	47267	Single-chamber implantable pacemaker, rate-responsive	CT141	Electrodes/Leads and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	14/09/2022	14/09/2022
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	14/09/2022	14/09/2022
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	14/09/2022	14/09/2022
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery	14/09/2022	14/09/2022
					CT145	Cardiovascular devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	14/09/2022	14/09/2022
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	14/09/2022	14/09/2022
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	14/09/2022	14/09/2022
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery	14/09/2022	14/09/2022
					CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	14/09/2022	14/09/2022
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	14/09/2022	14/09/2022
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	14/09/2022	14/09/2022
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery	14/09/2022	14/09/2022
					CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	14/09/2022	14/09/2022
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	14/09/2022	14/09/2022
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	14/09/2022	14/09/2022
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery	14/09/2022	14/09/2022
					CT233	Surgical	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	14/09/2022	14/09/2022
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	14/09/2022	14/09/2022
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	14/09/2022	14/09/2022
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery	14/09/2022	14/09/2022
					CT2406	Implantable	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	14/09/2022	14/09/2022
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	14/09/2022	14/09/2022
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	14/09/2022	14/09/2022
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery	14/09/2022	14/09/2022
					CT2407	Active implantable	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	14/09/2022	14/09/2022
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	14/09/2022	14/09/2022
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	14/09/2022	14/09/2022
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery	14/09/2022	14/09/2022
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	14/09/2022	14/09/2022
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	14/09/2022	14/09/2022
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	14/09/2022	14/09/2022
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery	14/09/2022	14/09/2022
					CT2727	Active implant devices/ancillaries	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	14/09/2022	14/09/2022
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	14/09/2022	14/09/2022
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	14/09/2022	14/09/2022
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery	14/09/2022	14/09/2022
					CT2750	Implantable cardiac electrical stimulators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	14/09/2022	14/09/2022
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	14/09/2022	14/09/2022
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	14/09/2022	14/09/2022
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery	14/09/2022	14/09/2022
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	14/09/2022	14/09/2022
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	14/09/2022	14/09/2022
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	14/09/2022	14/09/2022
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery	14/09/2022	14/09/2022
					CT2985	Active	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	14/09/2022	14/09/2022
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	14/09/2022	14/09/2022
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	14/09/2022	14/09/2022
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery	14/09/2022	14/09/2022
					CT3001	Implantable metal	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	14/09/2022	14/09/2022
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	14/09/2022	14/09/2022
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	14/09/2022	14/09/2022
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery	14/09/2022	14/09/2022
					CT3002	Implantable synthetic polymer	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	14/09/2022	14/09/2022
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	14/09/2022	14/09/2022
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	14/09/2022	14/09/2022

						F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery	14/09/2022	14/09/2022
					CT3006	Implantable inorganic material	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	14/09/2022	14/09/2022
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	14/09/2022	14/09/2022
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	14/09/2022	14/09/2022
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery	14/09/2022	14/09/2022
					CT314	Electrically powered	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	14/09/2022	14/09/2022
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	14/09/2022	14/09/2022
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	14/09/2022	14/09/2022
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery	14/09/2022	14/09/2022
					CT321	Long-term surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	14/09/2022	14/09/2022
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	14/09/2022	14/09/2022
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	14/09/2022	14/09/2022
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery	14/09/2022	14/09/2022
					CT449	Generators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	14/09/2022	14/09/2022
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	14/09/2022	14/09/2022
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	14/09/2022	14/09/2022
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery	14/09/2022	14/09/2022
					CT981	Single use	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	14/09/2022	14/09/2022
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	14/09/2022	14/09/2022
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	14/09/2022	14/09/2022
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery	14/09/2022	14/09/2022
					CT983	Surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	14/09/2022	14/09/2022
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	14/09/2022	14/09/2022
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	14/09/2022	14/09/2022
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery	14/09/2022	14/09/2022
7712	2022	All other reportable incidents	47267	Single-chamber implantable pacemaker, rate-responsive	CT141	Electrodes/Leads and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT145	Cardiovascular devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT233	Surgical	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT2406	Implantable	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT2407	Active implantable	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT2727	Active implant devices/ancillaries	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT2750	Implantable cardiac electrical stimulators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT2985	Active	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT3001	Implantable metal	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT3002	Implantable synthetic polymer	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT3006	Implantable inorganic material	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT314	Electrically powered	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT321	Long-term surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT449	Generators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT981	Single use	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT983	Surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
7713	2022	All other reportable incidents	17882	Physiologic-monitoring defibrillation system	CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A1603	Failure of Device to Self-Test		
							F27	Problem Identified During Non-Clinical Procedure	A1603	Failure of Device to Self-Test		
7714	2022	All other reportable incidents	60789	Intracardiac pacemaker	CT130	Catheters and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A051201	Device Dislodged or Dislocated	24/01/2022	24/01/2022
									A070102	High Capture Threshold	24/01/2022	24/01/2022
							F26	No Health Consequences or Impact	A051201	Device Dislodged or Dislocated	24/01/2022	24/01/2022
									A070102	High Capture Threshold	24/01/2022	24/01/2022
					CT145	Cardiovascular devices	E2403	No Clinical Signs, Symptoms or Conditions	A051201	Device Dislodged or Dislocated	24/01/2022	24/01/2022
									A070102	High Capture Threshold	24/01/2022	24/01/2022
							F26	No Health Consequences or Impact	A051201	Device Dislodged or Dislocated	24/01/2022	24/01/2022
									A070102	High Capture Threshold	24/01/2022	24/01/2022
					CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A051201	Device Dislodged or Dislocated	24/01/2022	24/01/2022
									A070102	High Capture Threshold	24/01/2022	24/01/2022
							F26	No Health Consequences or Impact	A051201	Device Dislodged or Dislocated	24/01/2022	24/01/2022
									A070102	High Capture Threshold	24/01/2022	24/01/2022
					CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A051201	Device Dislodged or Dislocated	24/01/2022	24/01/2022
									A070102	High Capture Threshold	24/01/2022	24/01/2022
							F26	No Health Consequences or Impact	A051201	Device Dislodged or Dislocated	24/01/2022	24/01/2022
									A070102	High Capture Threshold	24/01/2022	24/01/2022
					CT233	Surgical	E2403	No Clinical Signs, Symptoms or Conditions	A051201	Device Dislodged or Dislocated	24/01/2022	24/01/2022
									A070102	High Capture Threshold	24/01/2022	24/01/2022
							F26	No Health Consequences or Impact	A051201	Device Dislodged or Dislocated	24/01/2022	24/01/2022
									A070102	High Capture Threshold	24/01/2022	24/01/2022
					CT2406	Implantable	E2403	No Clinical Signs, Symptoms or Conditions	A051201	Device Dislodged or Dislocated	24/01/2022	24/01/2022

									A070102	High Capture Threshold	24/01/2022	24/01/2022
							F26	No Health Consequences or Impact	A051201	Device Dislodged or Dislocated	24/01/2022	24/01/2022
									A070102	High Capture Threshold	24/01/2022	24/01/2022
					CT2407	Active implantable	E2403	No Clinical Signs, Symptoms or Conditions	A051201	Device Dislodged or Dislocated	24/01/2022	24/01/2022
									A070102	High Capture Threshold	24/01/2022	24/01/2022
							F26	No Health Consequences or Impact	A051201	Device Dislodged or Dislocated	24/01/2022	24/01/2022
									A070102	High Capture Threshold	24/01/2022	24/01/2022
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A051201	Device Dislodged or Dislocated	24/01/2022	24/01/2022
									A070102	High Capture Threshold	24/01/2022	24/01/2022
							F26	No Health Consequences or Impact	A051201	Device Dislodged or Dislocated	24/01/2022	24/01/2022
									A070102	High Capture Threshold	24/01/2022	24/01/2022
					CT2727	Active implant devices/ancillaries	E2403	No Clinical Signs, Symptoms or Conditions	A051201	Device Dislodged or Dislocated	24/01/2022	24/01/2022
									A070102	High Capture Threshold	24/01/2022	24/01/2022
							F26	No Health Consequences or Impact	A051201	Device Dislodged or Dislocated	24/01/2022	24/01/2022
									A070102	High Capture Threshold	24/01/2022	24/01/2022
					CT2750	Implantable cardiac electrical stimulators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A051201	Device Dislodged or Dislocated	24/01/2022	24/01/2022
									A070102	High Capture Threshold	24/01/2022	24/01/2022
							F26	No Health Consequences or Impact	A051201	Device Dislodged or Dislocated	24/01/2022	24/01/2022
									A070102	High Capture Threshold	24/01/2022	24/01/2022
					CT2818	Catheterization/Cannulation/Fluid path devices	E2403	No Clinical Signs, Symptoms or Conditions	A051201	Device Dislodged or Dislocated	24/01/2022	24/01/2022
									A070102	High Capture Threshold	24/01/2022	24/01/2022
							F26	No Health Consequences or Impact	A051201	Device Dislodged or Dislocated	24/01/2022	24/01/2022
									A070102	High Capture Threshold	24/01/2022	24/01/2022
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A051201	Device Dislodged or Dislocated	24/01/2022	24/01/2022
									A070102	High Capture Threshold	24/01/2022	24/01/2022
							F26	No Health Consequences or Impact	A051201	Device Dislodged or Dislocated	24/01/2022	24/01/2022
									A070102	High Capture Threshold	24/01/2022	24/01/2022
					CT2985	Active	E2403	No Clinical Signs, Symptoms or Conditions	A051201	Device Dislodged or Dislocated	24/01/2022	24/01/2022
									A070102	High Capture Threshold	24/01/2022	24/01/2022
							F26	No Health Consequences or Impact	A051201	Device Dislodged or Dislocated	24/01/2022	24/01/2022
									A070102	High Capture Threshold	24/01/2022	24/01/2022
					CT3001	Implantable metal	E2403	No Clinical Signs, Symptoms or Conditions	A051201	Device Dislodged or Dislocated	24/01/2022	24/01/2022
									A070102	High Capture Threshold	24/01/2022	24/01/2022
							F26	No Health Consequences or Impact	A051201	Device Dislodged or Dislocated	24/01/2022	24/01/2022
									A070102	High Capture Threshold	24/01/2022	24/01/2022
					CT3002	Implantable synthetic polymer	E2403	No Clinical Signs, Symptoms or Conditions	A051201	Device Dislodged or Dislocated	24/01/2022	24/01/2022
									A070102	High Capture Threshold	24/01/2022	24/01/2022
							F26	No Health Consequences or Impact	A051201	Device Dislodged or Dislocated	24/01/2022	24/01/2022
									A070102	High Capture Threshold	24/01/2022	24/01/2022
					CT3006	Implantable inorganic material	E2403	No Clinical Signs, Symptoms or Conditions	A051201	Device Dislodged or Dislocated	24/01/2022	24/01/2022
									A070102	High Capture Threshold	24/01/2022	24/01/2022
							F26	No Health Consequences or Impact	A051201	Device Dislodged or Dislocated	24/01/2022	24/01/2022
									A070102	High Capture Threshold	24/01/2022	24/01/2022
					CT314	Electrically powered	E2403	No Clinical Signs, Symptoms or Conditions	A051201	Device Dislodged or Dislocated	24/01/2022	24/01/2022
									A070102	High Capture Threshold	24/01/2022	24/01/2022
							F26	No Health Consequences or Impact	A051201	Device Dislodged or Dislocated	24/01/2022	24/01/2022
									A070102	High Capture Threshold	24/01/2022	24/01/2022
					CT321	Long-term surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A051201	Device Dislodged or Dislocated	24/01/2022	24/01/2022
									A070102	High Capture Threshold	24/01/2022	24/01/2022
							F26	No Health Consequences or Impact	A051201	Device Dislodged or Dislocated	24/01/2022	24/01/2022
									A070102	High Capture Threshold	24/01/2022	24/01/2022
					CT981	Single use	E2403	No Clinical Signs, Symptoms or Conditions	A051201	Device Dislodged or Dislocated	24/01/2022	24/01/2022
									A070102	High Capture Threshold	24/01/2022	24/01/2022
							F26	No Health Consequences or Impact	A051201	Device Dislodged or Dislocated	24/01/2022	24/01/2022
									A070102	High Capture Threshold	24/01/2022	24/01/2022
					CT983	Surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A051201	Device Dislodged or Dislocated	24/01/2022	24/01/2022
									A070102	High Capture Threshold	24/01/2022	24/01/2022
							F26	No Health Consequences or Impact	A051201	Device Dislodged or Dislocated	24/01/2022	24/01/2022
									A070102	High Capture Threshold	24/01/2022	24/01/2022
7715	2022	All other reportable incidents	36241	Pacemaker/defibrillator lead	CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A072201	High Impedance	16/08/2022	16/08/2022
							F1908	Prolonged Surgery	A072201	High Impedance	16/08/2022	16/08/2022
7716	2022	All other reportable incidents	35223	Endocardial/interventricular septal pacing lead	CT1684	Cardiac pacemakers and associated devices	E1906	Unspecified Infection	A24	No Apparent Device Problem	29/05/2015	29/05/2015
							E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	29/05/2015	29/05/2015
							F08	Hospitalization or Prolonged Hospitalization	A24	No Apparent Device Problem	29/05/2015	29/05/2015
							F12	Serious Injury/ Illness/ Impairment	A24	No Apparent Device Problem	29/05/2015	29/05/2015
							F1905	Device Revision or Replacement	A24	No Apparent Device Problem	29/05/2015	29/05/2015
7717	2022	All other reportable incidents	47265	Dual-chamber implantable pacemaker, rate-responsive	CT141	Electrodes/Leads and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	09/06/2022	09/06/2022
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	09/06/2022	09/06/2022
							F1903	Device Explantation	A070504	Premature Discharge of Battery	09/06/2022	09/06/2022
					CT145	Cardiovascular devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	09/06/2022	09/06/2022
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	09/06/2022	09/06/2022
							F1903	Device Explantation	A070504	Premature Discharge of Battery	09/06/2022	09/06/2022
					CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	09/06/2022	09/06/2022
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	09/06/2022	09/06/2022
							F1903	Device Explantation	A070504	Premature Discharge of Battery	09/06/2022	09/06/2022
					CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	09/06/2022	09/06/2022
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	09/06/2022	09/06/2022
							F1903	Device Explantation	A070504	Premature Discharge of Battery	09/06/2022	09/06/2022
					CT233	Surgical	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	09/06/2022	09/06/2022
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	09/06/2022	09/06/2022
							F1903	Device Explantation	A070504	Premature Discharge of Battery	09/06/2022	09/06/2022
					CT2406	Implantable	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	09/06/2022	09/06/2022
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	09/06/2022	09/06/2022
							F1903	Device Explantation	A070504	Premature Discharge of Battery	09/06/2022	09/06/2022
					CT2407	Active implantable	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	09/06/2022	09/06/2022
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	09/06/2022	09/06/2022
							F1903	Device Explantation	A070504	Premature Discharge of Battery	09/06/2022	09/06/2022
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	09/06/2022	09/06/2022

							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	09/06/2022	09/06/2022
							F1903	Device Explantation	A070504	Premature Discharge of Battery	09/06/2022	09/06/2022
					CT2727	Active implant devices/ancillaries	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	09/06/2022	09/06/2022
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	09/06/2022	09/06/2022
							F1903	Device Explantation	A070504	Premature Discharge of Battery	09/06/2022	09/06/2022
					CT2750	Implantable cardiac electrical stimulators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	09/06/2022	09/06/2022
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	09/06/2022	09/06/2022
							F1903	Device Explantation	A070504	Premature Discharge of Battery	09/06/2022	09/06/2022
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	09/06/2022	09/06/2022
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	09/06/2022	09/06/2022
							F1903	Device Explantation	A070504	Premature Discharge of Battery	09/06/2022	09/06/2022
					CT2985	Active	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	09/06/2022	09/06/2022
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	09/06/2022	09/06/2022
							F1903	Device Explantation	A070504	Premature Discharge of Battery	09/06/2022	09/06/2022
					CT3001	Implantable metal	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	09/06/2022	09/06/2022
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	09/06/2022	09/06/2022
							F1903	Device Explantation	A070504	Premature Discharge of Battery	09/06/2022	09/06/2022
					CT3002	Implantable synthetic polymer	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	09/06/2022	09/06/2022
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	09/06/2022	09/06/2022
							F1903	Device Explantation	A070504	Premature Discharge of Battery	09/06/2022	09/06/2022
					CT3006	Implantable inorganic material	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	09/06/2022	09/06/2022
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	09/06/2022	09/06/2022
							F1903	Device Explantation	A070504	Premature Discharge of Battery	09/06/2022	09/06/2022
					CT314	Electrically powered	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	09/06/2022	09/06/2022
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	09/06/2022	09/06/2022
							F1903	Device Explantation	A070504	Premature Discharge of Battery	09/06/2022	09/06/2022
					CT321	Long-term surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	09/06/2022	09/06/2022
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	09/06/2022	09/06/2022
							F1903	Device Explantation	A070504	Premature Discharge of Battery	09/06/2022	09/06/2022
					CT449	Generators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	09/06/2022	09/06/2022
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	09/06/2022	09/06/2022
							F1903	Device Explantation	A070504	Premature Discharge of Battery	09/06/2022	09/06/2022
					CT981	Single use	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	09/06/2022	09/06/2022
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	09/06/2022	09/06/2022
							F1903	Device Explantation	A070504	Premature Discharge of Battery	09/06/2022	09/06/2022
					CT983	Surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	09/06/2022	09/06/2022
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	09/06/2022	09/06/2022
							F1903	Device Explantation	A070504	Premature Discharge of Battery	09/06/2022	09/06/2022
7718	2022	All other reportable incidents	35039	Epicardial pacing lead	CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070909	Oversensing		
									A070910	Undersensing		
									A090801	Signal Artifact/Noise		
							F08	Hospitalization or Prolonged Hospitalization	A070909	Oversensing		
									A070910	Undersensing		
									A090801	Signal Artifact/Noise		
							F12	Serious Injury/ Illness/ Impairment	A070909	Oversensing		
									A070910	Undersensing		
									A090801	Signal Artifact/Noise		
							F1905	Device Revision or Replacement	A070909	Oversensing		
									A070910	Undersensing		
									A090801	Signal Artifact/Noise		
							F23	Unexpected Medical Intervention	A070909	Oversensing		
									A070910	Undersensing		
									A090801	Signal Artifact/Noise		
7719	2022	All other reportable incidents	35852	Single-chamber implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem	04/02/2022	04/02/2022
									A070504	Premature Discharge of Battery	04/02/2022	04/02/2022
							F08	Hospitalization or Prolonged Hospitalization	A0705	Battery Problem	04/02/2022	04/02/2022
									A070504	Premature Discharge of Battery	04/02/2022	04/02/2022
							F1905	Device Revision or Replacement	A0705	Battery Problem	04/02/2022	04/02/2022
									A070504	Premature Discharge of Battery	04/02/2022	04/02/2022
7720	2022	All other reportable incidents	35223	Endocardial/interventricular septal pacing lead	CT1684	Cardiac pacemakers and associated devices	E011902	Presyncope	A070909	Oversensing		
							E0601	Arrhythmia	A070909	Oversensing		
							E2403	No Clinical Signs, Symptoms or Conditions	A070909	Oversensing		
							F23	Unexpected Medical Intervention	A070909	Oversensing		
7721	2022	All other reportable incidents	47265	Dual-chamber implantable pacemaker, rate-responsive	CT141	Electrodes/Leads and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	08/09/2022	08/09/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	08/09/2022	08/09/2022
					CT145	Cardiovascular devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	08/09/2022	08/09/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	08/09/2022	08/09/2022
					CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	08/09/2022	08/09/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	08/09/2022	08/09/2022
					CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	08/09/2022	08/09/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	08/09/2022	08/09/2022
					CT233	Surgical	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	08/09/2022	08/09/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	08/09/2022	08/09/2022
					CT2406	Implantable	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	08/09/2022	08/09/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	08/09/2022	08/09/2022
					CT2407	Active implantable	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	08/09/2022	08/09/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	08/09/2022	08/09/2022
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	08/09/2022	08/09/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	08/09/2022	08/09/2022
					CT2727	Active implant devices/ancillaries	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	08/09/2022	08/09/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	08/09/2022	08/09/2022
					CT2750	Implantable cardiac electrical stimulators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	08/09/2022	08/09/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	08/09/2022	08/09/2022
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	08/09/2022	08/09/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	08/09/2022	08/09/2022
					CT2985	Active	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	08/09/2022	08/09/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	08/09/2022	08/09/2022
					CT3001	Implantable metal	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	08/09/2022	08/09/2022

							F1901	Additional Surgery	A24	No Apparent Device Problem	08/09/2022	08/09/2022
					CT3002	Implantable synthetic polymer	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	08/09/2022	08/09/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	08/09/2022	08/09/2022
					CT3006	Implantable inorganic material	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	08/09/2022	08/09/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	08/09/2022	08/09/2022
					CT314	Electrically powered	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	08/09/2022	08/09/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	08/09/2022	08/09/2022
					CT321	Long-term surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	08/09/2022	08/09/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	08/09/2022	08/09/2022
					CT449	Generators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	08/09/2022	08/09/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	08/09/2022	08/09/2022
					CT981	Single use	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	08/09/2022	08/09/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	08/09/2022	08/09/2022
					CT983	Surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	08/09/2022	08/09/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	08/09/2022	08/09/2022
7722	2022		47265	Dual-chamber implantable pacemaker, rate-responsive	CT141	Electrodes/Leads and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	10/09/2022	10/09/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	10/09/2022	10/09/2022
					CT145	Cardiovascular devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	10/09/2022	10/09/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	10/09/2022	10/09/2022
					CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	10/09/2022	10/09/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	10/09/2022	10/09/2022
					CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	10/09/2022	10/09/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	10/09/2022	10/09/2022
					CT233	Surgical	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	10/09/2022	10/09/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	10/09/2022	10/09/2022
					CT2406	Implantable	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	10/09/2022	10/09/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	10/09/2022	10/09/2022
					CT2407	Active implantable	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	10/09/2022	10/09/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	10/09/2022	10/09/2022
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	10/09/2022	10/09/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	10/09/2022	10/09/2022
					CT2727	Active implant devices/ancillaries	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	10/09/2022	10/09/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	10/09/2022	10/09/2022
					CT2750	Implantable cardiac electrical stimulators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	10/09/2022	10/09/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	10/09/2022	10/09/2022
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	10/09/2022	10/09/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	10/09/2022	10/09/2022
					CT2985	Active	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	10/09/2022	10/09/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	10/09/2022	10/09/2022
					CT3001	Implantable metal	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	10/09/2022	10/09/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	10/09/2022	10/09/2022
					CT3002	Implantable synthetic polymer	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	10/09/2022	10/09/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	10/09/2022	10/09/2022
					CT3006	Implantable inorganic material	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	10/09/2022	10/09/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	10/09/2022	10/09/2022
					CT314	Electrically powered	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	10/09/2022	10/09/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	10/09/2022	10/09/2022
					CT321	Long-term surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	10/09/2022	10/09/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	10/09/2022	10/09/2022
					CT449	Generators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	10/09/2022	10/09/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	10/09/2022	10/09/2022
					CT981	Single use	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	10/09/2022	10/09/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	10/09/2022	10/09/2022
					CT983	Surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	10/09/2022	10/09/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	10/09/2022	10/09/2022
7723	2022	All other reportable incidents	47265	Dual-chamber implantable pacemaker, rate-responsive	CT141	Electrodes/Leads and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	10/09/2022	10/09/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	10/09/2022	10/09/2022
					CT145	Cardiovascular devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	10/09/2022	10/09/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	10/09/2022	10/09/2022
					CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	10/09/2022	10/09/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	10/09/2022	10/09/2022
					CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	10/09/2022	10/09/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	10/09/2022	10/09/2022
					CT233	Surgical	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	10/09/2022	10/09/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	10/09/2022	10/09/2022
					CT2406	Implantable	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	10/09/2022	10/09/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	10/09/2022	10/09/2022
					CT2407	Active implantable	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	10/09/2022	10/09/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	10/09/2022	10/09/2022
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	10/09/2022	10/09/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	10/09/2022	10/09/2022
					CT2727	Active implant devices/ancillaries	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	10/09/2022	10/09/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	10/09/2022	10/09/2022
					CT2750	Implantable cardiac electrical stimulators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	10/09/2022	10/09/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	10/09/2022	10/09/2022
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	10/09/2022	10/09/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	10/09/2022	10/09/2022
					CT2985	Active	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	10/09/2022	10/09/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	10/09/2022	10/09/2022
					CT3001	Implantable metal	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	10/09/2022	10/09/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	10/09/2022	10/09/2022
					CT3002	Implantable synthetic polymer	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	10/09/2022	10/09/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	10/09/2022	10/09/2022
					CT3006	Implantable inorganic material	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	10/09/2022	10/09/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	10/09/2022	10/09/2022
					CT314	Electrically powered	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	10/09/2022	10/09/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	10/09/2022	10/09/2022
					CT321	Long-term surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	10/09/2022	10/09/2022

						F1901	Additional Surgery	A24	No Apparent Device Problem	10/09/2022	10/09/2022
				CT449	Generators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	10/09/2022	10/09/2022
						F1901	Additional Surgery	A24	No Apparent Device Problem	10/09/2022	10/09/2022
				CT981	Single use	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	10/09/2022	10/09/2022
						F1901	Additional Surgery	A24	No Apparent Device Problem	10/09/2022	10/09/2022
				CT983	Surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	10/09/2022	10/09/2022
						F1901	Additional Surgery	A24	No Apparent Device Problem	10/09/2022	10/09/2022
7724	2022	All other reportable incidents	47265	Dual-chamber implantable pacemaker, rate-responsive	CT141	Electrodes/Leads and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	10/09/2022	10/09/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	10/09/2022	10/09/2022
					CT145	Cardiovascular devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	10/09/2022	10/09/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	10/09/2022	10/09/2022
					CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	10/09/2022	10/09/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	10/09/2022	10/09/2022
					CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	10/09/2022	10/09/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	10/09/2022	10/09/2022
					CT233	Surgical	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	10/09/2022	10/09/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	10/09/2022	10/09/2022
					CT2406	Implantable	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	10/09/2022	10/09/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	10/09/2022	10/09/2022
					CT2407	Active implantable	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	10/09/2022	10/09/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	10/09/2022	10/09/2022
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	10/09/2022	10/09/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	10/09/2022	10/09/2022
					CT2727	Active implant devices/ancillaries	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	10/09/2022	10/09/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	10/09/2022	10/09/2022
					CT2750	Implantable cardiac electrical stimulators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	10/09/2022	10/09/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	10/09/2022	10/09/2022
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	10/09/2022	10/09/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	10/09/2022	10/09/2022
					CT2985	Active	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	10/09/2022	10/09/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	10/09/2022	10/09/2022
					CT3001	Implantable metal	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	10/09/2022	10/09/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	10/09/2022	10/09/2022
					CT3002	Implantable synthetic polymer	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	10/09/2022	10/09/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	10/09/2022	10/09/2022
					CT3006	Implantable inorganic material	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	10/09/2022	10/09/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	10/09/2022	10/09/2022
					CT314	Electrically powered	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	10/09/2022	10/09/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	10/09/2022	10/09/2022
					CT321	Long-term surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	10/09/2022	10/09/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	10/09/2022	10/09/2022
					CT449	Generators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	10/09/2022	10/09/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	10/09/2022	10/09/2022
					CT981	Single use	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	10/09/2022	10/09/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	10/09/2022	10/09/2022
					CT983	Surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	10/09/2022	10/09/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	10/09/2022	10/09/2022
7725	2022	All other reportable incidents	47265	Dual-chamber implantable pacemaker, rate-responsive	CT141	Electrodes/Leads and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	10/09/2022	10/09/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	10/09/2022	10/09/2022
					CT145	Cardiovascular devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	10/09/2022	10/09/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	10/09/2022	10/09/2022
					CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	10/09/2022	10/09/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	10/09/2022	10/09/2022
					CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	10/09/2022	10/09/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	10/09/2022	10/09/2022
					CT233	Surgical	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	10/09/2022	10/09/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	10/09/2022	10/09/2022
					CT2406	Implantable	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	10/09/2022	10/09/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	10/09/2022	10/09/2022
					CT2407	Active implantable	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	10/09/2022	10/09/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	10/09/2022	10/09/2022
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	10/09/2022	10/09/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	10/09/2022	10/09/2022
					CT2727	Active implant devices/ancillaries	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	10/09/2022	10/09/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	10/09/2022	10/09/2022
					CT2750	Implantable cardiac electrical stimulators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	10/09/2022	10/09/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	10/09/2022	10/09/2022
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	10/09/2022	10/09/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	10/09/2022	10/09/2022
					CT2985	Active	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	10/09/2022	10/09/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	10/09/2022	10/09/2022
					CT3001	Implantable metal	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	10/09/2022	10/09/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	10/09/2022	10/09/2022
					CT3002	Implantable synthetic polymer	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	10/09/2022	10/09/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	10/09/2022	10/09/2022
					CT3006	Implantable inorganic material	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	10/09/2022	10/09/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	10/09/2022	10/09/2022
					CT314	Electrically powered	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	10/09/2022	10/09/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	10/09/2022	10/09/2022
					CT321	Long-term surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	10/09/2022	10/09/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	10/09/2022	10/09/2022
					CT449	Generators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	10/09/2022	10/09/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	10/09/2022	10/09/2022
					CT981	Single use	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	10/09/2022	10/09/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	10/09/2022	10/09/2022
					CT983	Surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	10/09/2022	10/09/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	10/09/2022	10/09/2022
7726	2022	All other reportable incidents	35223	Endocardial/interventricular septal pacing lead	CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A0401	Break	09/09/2022	09/09/2022

									A090801	Signal Artifact/Noise	09/09/2022	09/09/2022
							F08	Hospitalization or Prolonged Hospitalization	A0401	Break	09/09/2022	09/09/2022
									A090801	Signal Artifact/Noise	09/09/2022	09/09/2022
							F12	Serious Injury/ Illness/ Impairment	A0401	Break	09/09/2022	09/09/2022
									A090801	Signal Artifact/Noise	09/09/2022	09/09/2022
							F1905	Device Revision or Replacement	A0401	Break	09/09/2022	09/09/2022
									A090801	Signal Artifact/Noise	09/09/2022	09/09/2022
							F23	Unexpected Medical Intervention	A0401	Break	09/09/2022	09/09/2022
									A090801	Signal Artifact/Noise	09/09/2022	09/09/2022
7727	2022	All other reportable incidents	35852	Single-chamber implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A070909	Oversensing		
							F26	No Health Consequences or Impact	A070909	Oversensing		
7728	2022	All other reportable incidents	47265	Dual-chamber implantable pacemaker, rate-responsive	CT141	Electrodes/Leads and associated devices	E020204	Malaise	A0203	Defective Device		
							E0623	Unspecified Heart Problem	A0203	Defective Device		
							E0717	Dyspnea	A0203	Defective Device		
							E2103	Device Overstimulation of Tissue	A0203	Defective Device		
							E2312	Fatigue	A0203	Defective Device		
							E2403	No Clinical Signs, Symptoms or Conditions	A0203	Defective Device		
							F15	Recognized Device or Procedural Complication	A0203	Defective Device		
							F2203	Imaging Required	A0203	Defective Device		
					CT145	Cardiovascular devices	E020204	Malaise	A0203	Defective Device		
							E0623	Unspecified Heart Problem	A0203	Defective Device		
							E0717	Dyspnea	A0203	Defective Device		
							E2103	Device Overstimulation of Tissue	A0203	Defective Device		
							E2312	Fatigue	A0203	Defective Device		
							E2403	No Clinical Signs, Symptoms or Conditions	A0203	Defective Device		
							F15	Recognized Device or Procedural Complication	A0203	Defective Device		
							F2203	Imaging Required	A0203	Defective Device		
					CT1684	Cardiac pacemakers and associated devices	E020204	Malaise	A0203	Defective Device		
							E0623	Unspecified Heart Problem	A0203	Defective Device		
							E0717	Dyspnea	A0203	Defective Device		
							E2103	Device Overstimulation of Tissue	A0203	Defective Device		
							E2312	Fatigue	A0203	Defective Device		
							E2403	No Clinical Signs, Symptoms or Conditions	A0203	Defective Device		
							F15	Recognized Device or Procedural Complication	A0203	Defective Device		
							F2203	Imaging Required	A0203	Defective Device		
					CT1685	Cardiac pacemakers	E020204	Malaise	A0203	Defective Device		
							E0623	Unspecified Heart Problem	A0203	Defective Device		
							E0717	Dyspnea	A0203	Defective Device		
							E2103	Device Overstimulation of Tissue	A0203	Defective Device		
							E2312	Fatigue	A0203	Defective Device		
							E2403	No Clinical Signs, Symptoms or Conditions	A0203	Defective Device		
							F15	Recognized Device or Procedural Complication	A0203	Defective Device		
							F2203	Imaging Required	A0203	Defective Device		
					CT233	Surgical	E020204	Malaise	A0203	Defective Device		
							E0623	Unspecified Heart Problem	A0203	Defective Device		
							E0717	Dyspnea	A0203	Defective Device		
							E2103	Device Overstimulation of Tissue	A0203	Defective Device		
							E2312	Fatigue	A0203	Defective Device		
							E2403	No Clinical Signs, Symptoms or Conditions	A0203	Defective Device		
							F15	Recognized Device or Procedural Complication	A0203	Defective Device		
							F2203	Imaging Required	A0203	Defective Device		
					CT2406	Implantable	E020204	Malaise	A0203	Defective Device		
							E0623	Unspecified Heart Problem	A0203	Defective Device		
							E0717	Dyspnea	A0203	Defective Device		
							E2103	Device Overstimulation of Tissue	A0203	Defective Device		
							E2312	Fatigue	A0203	Defective Device		
							E2403	No Clinical Signs, Symptoms or Conditions	A0203	Defective Device		
							F15	Recognized Device or Procedural Complication	A0203	Defective Device		
							F2203	Imaging Required	A0203	Defective Device		
					CT2407	Active implantable	E020204	Malaise	A0203	Defective Device		
							E0623	Unspecified Heart Problem	A0203	Defective Device		
							E0717	Dyspnea	A0203	Defective Device		
							E2103	Device Overstimulation of Tissue	A0203	Defective Device		
							E2312	Fatigue	A0203	Defective Device		
							E2403	No Clinical Signs, Symptoms or Conditions	A0203	Defective Device		
							F15	Recognized Device or Procedural Complication	A0203	Defective Device		
							F2203	Imaging Required	A0203	Defective Device		
					CT257	Implantable cardiac pacemakers	E020204	Malaise	A0203	Defective Device		
							E0623	Unspecified Heart Problem	A0203	Defective Device		
							E0717	Dyspnea	A0203	Defective Device		
							E2103	Device Overstimulation of Tissue	A0203	Defective Device		
							E2312	Fatigue	A0203	Defective Device		
							E2403	No Clinical Signs, Symptoms or Conditions	A0203	Defective Device		
							F15	Recognized Device or Procedural Complication	A0203	Defective Device		
							F2203	Imaging Required	A0203	Defective Device		
					CT2727	Active implant devices/ancillaries	E020204	Malaise	A0203	Defective Device		
							E0623	Unspecified Heart Problem	A0203	Defective Device		
							E0717	Dyspnea	A0203	Defective Device		
							E2103	Device Overstimulation of Tissue	A0203	Defective Device		
							E2312	Fatigue	A0203	Defective Device		
							E2403	No Clinical Signs, Symptoms or Conditions	A0203	Defective Device		
							F15	Recognized Device or Procedural Complication	A0203	Defective Device		
							F2203	Imaging Required	A0203	Defective Device		
					CT2750	Implantable cardiac electrical stimulators and associated devices	E020204	Malaise	A0203	Defective Device		
							E0623	Unspecified Heart Problem	A0203	Defective Device		
							E0717	Dyspnea	A0203	Defective Device		
							E2103	Device Overstimulation of Tissue	A0203	Defective Device		
							E2312	Fatigue	A0203	Defective Device		

							E2403	No Clinical Signs, Symptoms or Conditions	A0203	Defective Device		
							F15	Recognized Device or Procedural Complication	A0203	Defective Device		
							F2203	Imaging Required	A0203	Defective Device		
					CT2866	Implantable cardiac electrical stimulators	E020204	Malaise	A0203	Defective Device		
							E0623	Unspecified Heart Problem	A0203	Defective Device		
							E0717	Dyspnea	A0203	Defective Device		
							E2103	Device Overstimulation of Tissue	A0203	Defective Device		
							E2312	Fatigue	A0203	Defective Device		
							E2403	No Clinical Signs, Symptoms or Conditions	A0203	Defective Device		
							F15	Recognized Device or Procedural Complication	A0203	Defective Device		
							F2203	Imaging Required	A0203	Defective Device		
					CT2985	Active	E020204	Malaise	A0203	Defective Device		
							E0623	Unspecified Heart Problem	A0203	Defective Device		
							E0717	Dyspnea	A0203	Defective Device		
							E2103	Device Overstimulation of Tissue	A0203	Defective Device		
							E2312	Fatigue	A0203	Defective Device		
							E2403	No Clinical Signs, Symptoms or Conditions	A0203	Defective Device		
							F15	Recognized Device or Procedural Complication	A0203	Defective Device		
							F2203	Imaging Required	A0203	Defective Device		
					CT3001	Implantable metal	E020204	Malaise	A0203	Defective Device		
							E0623	Unspecified Heart Problem	A0203	Defective Device		
							E0717	Dyspnea	A0203	Defective Device		
							E2103	Device Overstimulation of Tissue	A0203	Defective Device		
							E2312	Fatigue	A0203	Defective Device		
							E2403	No Clinical Signs, Symptoms or Conditions	A0203	Defective Device		
							F15	Recognized Device or Procedural Complication	A0203	Defective Device		
							F2203	Imaging Required	A0203	Defective Device		
					CT3002	Implantable synthetic polymer	E020204	Malaise	A0203	Defective Device		
							E0623	Unspecified Heart Problem	A0203	Defective Device		
							E0717	Dyspnea	A0203	Defective Device		
							E2103	Device Overstimulation of Tissue	A0203	Defective Device		
							E2312	Fatigue	A0203	Defective Device		
							E2403	No Clinical Signs, Symptoms or Conditions	A0203	Defective Device		
							F15	Recognized Device or Procedural Complication	A0203	Defective Device		
							F2203	Imaging Required	A0203	Defective Device		
					CT3006	Implantable inorganic material	E020204	Malaise	A0203	Defective Device		
							E0623	Unspecified Heart Problem	A0203	Defective Device		
							E0717	Dyspnea	A0203	Defective Device		
							E2103	Device Overstimulation of Tissue	A0203	Defective Device		
							E2312	Fatigue	A0203	Defective Device		
							E2403	No Clinical Signs, Symptoms or Conditions	A0203	Defective Device		
							F15	Recognized Device or Procedural Complication	A0203	Defective Device		
							F2203	Imaging Required	A0203	Defective Device		
					CT314	Electrically powered	E020204	Malaise	A0203	Defective Device		
							E0623	Unspecified Heart Problem	A0203	Defective Device		
							E0717	Dyspnea	A0203	Defective Device		
							E2103	Device Overstimulation of Tissue	A0203	Defective Device		
							E2312	Fatigue	A0203	Defective Device		
							E2403	No Clinical Signs, Symptoms or Conditions	A0203	Defective Device		
							F15	Recognized Device or Procedural Complication	A0203	Defective Device		
							F2203	Imaging Required	A0203	Defective Device		
					CT321	Long-term surgical invasive	E020204	Malaise	A0203	Defective Device		
							E0623	Unspecified Heart Problem	A0203	Defective Device		
							E0717	Dyspnea	A0203	Defective Device		
							E2103	Device Overstimulation of Tissue	A0203	Defective Device		
							E2312	Fatigue	A0203	Defective Device		
							E2403	No Clinical Signs, Symptoms or Conditions	A0203	Defective Device		
							F15	Recognized Device or Procedural Complication	A0203	Defective Device		
							F2203	Imaging Required	A0203	Defective Device		
					CT449	Generators and associated devices	E020204	Malaise	A0203	Defective Device		
							E0623	Unspecified Heart Problem	A0203	Defective Device		
							E0717	Dyspnea	A0203	Defective Device		
							E2103	Device Overstimulation of Tissue	A0203	Defective Device		
							E2312	Fatigue	A0203	Defective Device		
							E2403	No Clinical Signs, Symptoms or Conditions	A0203	Defective Device		
							F15	Recognized Device or Procedural Complication	A0203	Defective Device		
							F2203	Imaging Required	A0203	Defective Device		
					CT981	Single use	E020204	Malaise	A0203	Defective Device		
							E0623	Unspecified Heart Problem	A0203	Defective Device		
							E0717	Dyspnea	A0203	Defective Device		
							E2103	Device Overstimulation of Tissue	A0203	Defective Device		
							E2312	Fatigue	A0203	Defective Device		
							E2403	No Clinical Signs, Symptoms or Conditions	A0203	Defective Device		
							F15	Recognized Device or Procedural Complication	A0203	Defective Device		
							F2203	Imaging Required	A0203	Defective Device		
					CT983	Surgical invasive	E020204	Malaise	A0203	Defective Device		
							E0623	Unspecified Heart Problem	A0203	Defective Device		
							E0717	Dyspnea	A0203	Defective Device		
							E2103	Device Overstimulation of Tissue	A0203	Defective Device		
							E2312	Fatigue	A0203	Defective Device		
							E2403	No Clinical Signs, Symptoms or Conditions	A0203	Defective Device		
							F15	Recognized Device or Procedural Complication	A0203	Defective Device		
							F2203	Imaging Required	A0203	Defective Device		
7729	2022	All other reportable incidents	35223	Endocardial/interventricular septal pacing lead	CT1684	Cardiac pacemakers and associated devices	E2401	Insufficient Information	A070909	Oversensing		
									A070910	Undersensing		
							E2403	No Clinical Signs, Symptoms or Conditions	A070909	Oversensing		
									A070910	Undersensing		
							F24	Insufficient Health Impact Information	A070909	Oversensing		

									A070910	Undersensing		
							F26	No Health Consequences or Impact	A070909	Oversensing		
									A070910	Undersensing		
7730	2022	All other reportable incidents	35223	Endocardial/interventricular septal pacing lead	CT1684	Cardiac pacemakers and associated devices	E2114	Perforation	A24	No Apparent Device Problem		
							E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem		
							F08	Hospitalization or Prolonged Hospitalization	A24	No Apparent Device Problem		
							F1203	Life Threatening Illness or Injury	A24	No Apparent Device Problem		
							F1905	Device Revision or Replacement	A24	No Apparent Device Problem		
7731	2022	All other reportable incidents	47265	Dual-chamber implantable pacemaker, rate-responsive	CT141	Electrodes/Leads and associated devices	E011903	Syncope/Fainting	A070504	Premature Discharge of Battery	28/01/2022	28/01/2022
									A071102	Failure to Interrogate	28/01/2022	28/01/2022
									A0712	Pacing Problem	28/01/2022	28/01/2022
							E060106	Heart Block	A070504	Premature Discharge of Battery	28/01/2022	28/01/2022
									A071102	Failure to Interrogate	28/01/2022	28/01/2022
									A0712	Pacing Problem	28/01/2022	28/01/2022
							E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	28/01/2022	28/01/2022
									A071102	Failure to Interrogate	28/01/2022	28/01/2022
									A0712	Pacing Problem	28/01/2022	28/01/2022
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	28/01/2022	28/01/2022
									A071102	Failure to Interrogate	28/01/2022	28/01/2022
									A0712	Pacing Problem	28/01/2022	28/01/2022
							F1203	Life Threatening Illness or Injury	A070504	Premature Discharge of Battery	28/01/2022	28/01/2022
									A071102	Failure to Interrogate	28/01/2022	28/01/2022
									A0712	Pacing Problem	28/01/2022	28/01/2022
					CT145	Cardiovascular devices	E011903	Syncope/Fainting	A070504	Premature Discharge of Battery	28/01/2022	28/01/2022
									A071102	Failure to Interrogate	28/01/2022	28/01/2022
									A0712	Pacing Problem	28/01/2022	28/01/2022
							E060106	Heart Block	A070504	Premature Discharge of Battery	28/01/2022	28/01/2022
									A071102	Failure to Interrogate	28/01/2022	28/01/2022
									A0712	Pacing Problem	28/01/2022	28/01/2022
							E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	28/01/2022	28/01/2022
									A071102	Failure to Interrogate	28/01/2022	28/01/2022
									A0712	Pacing Problem	28/01/2022	28/01/2022
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	28/01/2022	28/01/2022
									A071102	Failure to Interrogate	28/01/2022	28/01/2022
									A0712	Pacing Problem	28/01/2022	28/01/2022
							F1203	Life Threatening Illness or Injury	A070504	Premature Discharge of Battery	28/01/2022	28/01/2022
									A071102	Failure to Interrogate	28/01/2022	28/01/2022
									A0712	Pacing Problem	28/01/2022	28/01/2022
					CT1684	Cardiac pacemakers and associated devices	E011903	Syncope/Fainting	A070504	Premature Discharge of Battery	28/01/2022	28/01/2022
									A071102	Failure to Interrogate	28/01/2022	28/01/2022
									A0712	Pacing Problem	28/01/2022	28/01/2022
							E060106	Heart Block	A070504	Premature Discharge of Battery	28/01/2022	28/01/2022
									A071102	Failure to Interrogate	28/01/2022	28/01/2022
									A0712	Pacing Problem	28/01/2022	28/01/2022
							E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	28/01/2022	28/01/2022
									A071102	Failure to Interrogate	28/01/2022	28/01/2022
									A0712	Pacing Problem	28/01/2022	28/01/2022
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	28/01/2022	28/01/2022
									A071102	Failure to Interrogate	28/01/2022	28/01/2022
									A0712	Pacing Problem	28/01/2022	28/01/2022
							F1203	Life Threatening Illness or Injury	A070504	Premature Discharge of Battery	28/01/2022	28/01/2022
									A071102	Failure to Interrogate	28/01/2022	28/01/2022
									A0712	Pacing Problem	28/01/2022	28/01/2022
					CT1685	Cardiac pacemakers	E011903	Syncope/Fainting	A070504	Premature Discharge of Battery	28/01/2022	28/01/2022
									A071102	Failure to Interrogate	28/01/2022	28/01/2022
									A0712	Pacing Problem	28/01/2022	28/01/2022
							E060106	Heart Block	A070504	Premature Discharge of Battery	28/01/2022	28/01/2022
									A071102	Failure to Interrogate	28/01/2022	28/01/2022
									A0712	Pacing Problem	28/01/2022	28/01/2022
							E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	28/01/2022	28/01/2022
									A071102	Failure to Interrogate	28/01/2022	28/01/2022
									A0712	Pacing Problem	28/01/2022	28/01/2022
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	28/01/2022	28/01/2022
									A071102	Failure to Interrogate	28/01/2022	28/01/2022
									A0712	Pacing Problem	28/01/2022	28/01/2022
							F1203	Life Threatening Illness or Injury	A070504	Premature Discharge of Battery	28/01/2022	28/01/2022
									A071102	Failure to Interrogate	28/01/2022	28/01/2022
									A0712	Pacing Problem	28/01/2022	28/01/2022
					CT233	Surgical	E011903	Syncope/Fainting	A070504	Premature Discharge of Battery	28/01/2022	28/01/2022
									A071102	Failure to Interrogate	28/01/2022	28/01/2022
									A0712	Pacing Problem	28/01/2022	28/01/2022
							E060106	Heart Block	A070504	Premature Discharge of Battery	28/01/2022	28/01/2022
									A071102	Failure to Interrogate	28/01/2022	28/01/2022
									A0712	Pacing Problem	28/01/2022	28/01/2022
							E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	28/01/2022	28/01/2022
									A071102	Failure to Interrogate	28/01/2022	28/01/2022
									A0712	Pacing Problem	28/01/2022	28/01/2022
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	28/01/2022	28/01/2022
									A071102	Failure to Interrogate	28/01/2022	28/01/2022
									A0712	Pacing Problem	28/01/2022	28/01/2022
							F1203	Life Threatening Illness or Injury	A070504	Premature Discharge of Battery	28/01/2022	28/01/2022
									A071102	Failure to Interrogate	28/01/2022	28/01/2022
									A0712	Pacing Problem	28/01/2022	28/01/2022
					CT2406	Implantable	E011903	Syncope/Fainting	A070504	Premature Discharge of Battery	28/01/2022	28/01/2022
									A071102	Failure to Interrogate	28/01/2022	28/01/2022
									A0712	Pacing Problem	28/01/2022	28/01/2022
							E060106	Heart Block	A070504	Premature Discharge of Battery	28/01/2022	28/01/2022
									A071102	Failure to Interrogate	28/01/2022	28/01/2022

								A0712	Pacing Problem	28/01/2022	28/01/2022
						E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	28/01/2022	28/01/2022
								A071102	Failure to Interrogate	28/01/2022	28/01/2022
								A0712	Pacing Problem	28/01/2022	28/01/2022
						F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	28/01/2022	28/01/2022
								A071102	Failure to Interrogate	28/01/2022	28/01/2022
								A0712	Pacing Problem	28/01/2022	28/01/2022
						F1203	Life Threatening Illness or Injury	A070504	Premature Discharge of Battery	28/01/2022	28/01/2022
								A071102	Failure to Interrogate	28/01/2022	28/01/2022
								A0712	Pacing Problem	28/01/2022	28/01/2022
					CT2407	Active implantable	E011903	Syncope/Fainting	A070504	Premature Discharge of Battery	28/01/2022	28/01/2022
									A071102	Failure to Interrogate	28/01/2022	28/01/2022
									A0712	Pacing Problem	28/01/2022	28/01/2022
						E060106	Heart Block	A070504	Premature Discharge of Battery	28/01/2022	28/01/2022
								A071102	Failure to Interrogate	28/01/2022	28/01/2022
								A0712	Pacing Problem	28/01/2022	28/01/2022
						E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	28/01/2022	28/01/2022
								A071102	Failure to Interrogate	28/01/2022	28/01/2022
								A0712	Pacing Problem	28/01/2022	28/01/2022
						F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	28/01/2022	28/01/2022
								A071102	Failure to Interrogate	28/01/2022	28/01/2022
								A0712	Pacing Problem	28/01/2022	28/01/2022
						F1203	Life Threatening Illness or Injury	A070504	Premature Discharge of Battery	28/01/2022	28/01/2022
								A071102	Failure to Interrogate	28/01/2022	28/01/2022
								A0712	Pacing Problem	28/01/2022	28/01/2022
					CT257	Implantable cardiac pacemakers	E011903	Syncope/Fainting	A070504	Premature Discharge of Battery	28/01/2022	28/01/2022
									A071102	Failure to Interrogate	28/01/2022	28/01/2022
									A0712	Pacing Problem	28/01/2022	28/01/2022
						E060106	Heart Block	A070504	Premature Discharge of Battery	28/01/2022	28/01/2022
								A071102	Failure to Interrogate	28/01/2022	28/01/2022
								A0712	Pacing Problem	28/01/2022	28/01/2022
						E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	28/01/2022	28/01/2022
								A071102	Failure to Interrogate	28/01/2022	28/01/2022
								A0712	Pacing Problem	28/01/2022	28/01/2022
						F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	28/01/2022	28/01/2022
								A071102	Failure to Interrogate	28/01/2022	28/01/2022
								A0712	Pacing Problem	28/01/2022	28/01/2022
						F1203	Life Threatening Illness or Injury	A070504	Premature Discharge of Battery	28/01/2022	28/01/2022
								A071102	Failure to Interrogate	28/01/2022	28/01/2022
								A0712	Pacing Problem	28/01/2022	28/01/2022
					CT2727	Active implant devices/ancillaries	E011903	Syncope/Fainting	A070504	Premature Discharge of Battery	28/01/2022	28/01/2022
									A071102	Failure to Interrogate	28/01/2022	28/01/2022
									A0712	Pacing Problem	28/01/2022	28/01/2022
						E060106	Heart Block	A070504	Premature Discharge of Battery	28/01/2022	28/01/2022
								A071102	Failure to Interrogate	28/01/2022	28/01/2022
								A0712	Pacing Problem	28/01/2022	28/01/2022
						E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	28/01/2022	28/01/2022
								A071102	Failure to Interrogate	28/01/2022	28/01/2022
								A0712	Pacing Problem	28/01/2022	28/01/2022
						F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	28/01/2022	28/01/2022
								A071102	Failure to Interrogate	28/01/2022	28/01/2022
								A0712	Pacing Problem	28/01/2022	28/01/2022
						F1203	Life Threatening Illness or Injury	A070504	Premature Discharge of Battery	28/01/2022	28/01/2022
								A071102	Failure to Interrogate	28/01/2022	28/01/2022
								A0712	Pacing Problem	28/01/2022	28/01/2022
					CT2750	Implantable cardiac electrical stimulators and associated devices	E011903	Syncope/Fainting	A070504	Premature Discharge of Battery	28/01/2022	28/01/2022
									A071102	Failure to Interrogate	28/01/2022	28/01/2022
									A0712	Pacing Problem	28/01/2022	28/01/2022
						E060106	Heart Block	A070504	Premature Discharge of Battery	28/01/2022	28/01/2022
								A071102	Failure to Interrogate	28/01/2022	28/01/2022
								A0712	Pacing Problem	28/01/2022	28/01/2022
						E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	28/01/2022	28/01/2022
								A071102	Failure to Interrogate	28/01/2022	28/01/2022
								A0712	Pacing Problem	28/01/2022	28/01/2022
						F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	28/01/2022	28/01/2022
								A071102	Failure to Interrogate	28/01/2022	28/01/2022
								A0712	Pacing Problem	28/01/2022	28/01/2022
						F1203	Life Threatening Illness or Injury	A070504	Premature Discharge of Battery	28/01/2022	28/01/2022
								A071102	Failure to Interrogate	28/01/2022	28/01/2022
								A0712	Pacing Problem	28/01/2022	28/01/2022
					CT2866	Implantable cardiac electrical stimulators	E011903	Syncope/Fainting	A070504	Premature Discharge of Battery	28/01/2022	28/01/2022
									A071102	Failure to Interrogate	28/01/2022	28/01/2022
									A0712	Pacing Problem	28/01/2022	28/01/2022
						E060106	Heart Block	A070504	Premature Discharge of Battery	28/01/2022	28/01/2022
								A071102	Failure to Interrogate	28/01/2022	28/01/2022
								A0712	Pacing Problem	28/01/2022	28/01/2022
						E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	28/01/2022	28/01/2022
								A071102	Failure to Interrogate	28/01/2022	28/01/2022
								A0712	Pacing Problem	28/01/2022	28/01/2022
						F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	28/01/2022	28/01/2022
								A071102	Failure to Interrogate	28/01/2022	28/01/2022
								A0712	Pacing Problem	28/01/2022	28/01/2022
						F1203	Life Threatening Illness or Injury	A070504	Premature Discharge of Battery	28/01/2022	28/01/2022
								A071102	Failure to Interrogate	28/01/2022	28/01/2022
								A0712	Pacing Problem	28/01/2022	28/01/2022
					CT2985	Active	E011903	Syncope/Fainting	A070504	Premature Discharge of Battery	28/01/2022	28/01/2022
									A071102	Failure to Interrogate	28/01/2022	28/01/2022
									A0712	Pacing Problem	28/01/2022	28/01/2022

							E060106	Heart Block	A070504	Premature Discharge of Battery	28/01/2022	28/01/2022
									A071102	Failure to Interrogate	28/01/2022	28/01/2022
									A0712	Pacing Problem	28/01/2022	28/01/2022
							E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	28/01/2022	28/01/2022
									A071102	Failure to Interrogate	28/01/2022	28/01/2022
									A0712	Pacing Problem	28/01/2022	28/01/2022
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	28/01/2022	28/01/2022
									A071102	Failure to Interrogate	28/01/2022	28/01/2022
									A0712	Pacing Problem	28/01/2022	28/01/2022
							F1203	Life Threatening Illness or Injury	A070504	Premature Discharge of Battery	28/01/2022	28/01/2022
									A071102	Failure to Interrogate	28/01/2022	28/01/2022
									A0712	Pacing Problem	28/01/2022	28/01/2022
						CT3001	Implantable metal	E011903	Syncope/Fainting	A070504	Premature Discharge of Battery	28/01/2022	28/01/2022
										A071102	Failure to Interrogate	28/01/2022	28/01/2022
										A0712	Pacing Problem	28/01/2022	28/01/2022
								E060106	Heart Block	A070504	Premature Discharge of Battery	28/01/2022	28/01/2022
										A071102	Failure to Interrogate	28/01/2022	28/01/2022
										A0712	Pacing Problem	28/01/2022	28/01/2022
								E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	28/01/2022	28/01/2022
										A071102	Failure to Interrogate	28/01/2022	28/01/2022
										A0712	Pacing Problem	28/01/2022	28/01/2022
								F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	28/01/2022	28/01/2022
										A071102	Failure to Interrogate	28/01/2022	28/01/2022
										A0712	Pacing Problem	28/01/2022	28/01/2022
								F1203	Life Threatening Illness or Injury	A070504	Premature Discharge of Battery	28/01/2022	28/01/2022
										A071102	Failure to Interrogate	28/01/2022	28/01/2022
										A0712	Pacing Problem	28/01/2022	28/01/2022
						CT3002	Implantable synthetic polymer	E011903	Syncope/Fainting	A070504	Premature Discharge of Battery	28/01/2022	28/01/2022
										A071102	Failure to Interrogate	28/01/2022	28/01/2022
										A0712	Pacing Problem	28/01/2022	28/01/2022
								E060106	Heart Block	A070504	Premature Discharge of Battery	28/01/2022	28/01/2022
										A071102	Failure to Interrogate	28/01/2022	28/01/2022
										A0712	Pacing Problem	28/01/2022	28/01/2022
								E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	28/01/2022	28/01/2022
										A071102	Failure to Interrogate	28/01/2022	28/01/2022
										A0712	Pacing Problem	28/01/2022	28/01/2022
								F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	28/01/2022	28/01/2022
										A071102	Failure to Interrogate	28/01/2022	28/01/2022
										A0712	Pacing Problem	28/01/2022	28/01/2022
								F1203	Life Threatening Illness or Injury	A070504	Premature Discharge of Battery	28/01/2022	28/01/2022
										A071102	Failure to Interrogate	28/01/2022	28/01/2022
										A0712	Pacing Problem	28/01/2022	28/01/2022
						CT3006	Implantable inorganic material	E011903	Syncope/Fainting	A070504	Premature Discharge of Battery	28/01/2022	28/01/2022
										A071102	Failure to Interrogate	28/01/2022	28/01/2022
										A0712	Pacing Problem	28/01/2022	28/01/2022
								E060106	Heart Block	A070504	Premature Discharge of Battery	28/01/2022	28/01/2022
										A071102	Failure to Interrogate	28/01/2022	28/01/2022
										A0712	Pacing Problem	28/01/2022	28/01/2022
								E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	28/01/2022	28/01/2022
										A071102	Failure to Interrogate	28/01/2022	28/01/2022
										A0712	Pacing Problem	28/01/2022	28/01/2022
								F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	28/01/2022	28/01/2022
										A071102	Failure to Interrogate	28/01/2022	28/01/2022
										A0712	Pacing Problem	28/01/2022	28/01/2022
								F1203	Life Threatening Illness or Injury	A070504	Premature Discharge of Battery	28/01/2022	28/01/2022
										A071102	Failure to Interrogate	28/01/2022	28/01/2022
										A0712	Pacing Problem	28/01/2022	28/01/2022
						CT314	Electrically powered	E011903	Syncope/Fainting	A070504	Premature Discharge of Battery	28/01/2022	28/01/2022
										A071102	Failure to Interrogate	28/01/2022	28/01/2022
										A0712	Pacing Problem	28/01/2022	28/01/2022
								E060106	Heart Block	A070504	Premature Discharge of Battery	28/01/2022	28/01/2022
										A071102	Failure to Interrogate	28/01/2022	28/01/2022
										A0712	Pacing Problem	28/01/2022	28/01/2022
								E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	28/01/2022	28/01/2022
										A071102	Failure to Interrogate	28/01/2022	28/01/2022
										A0712	Pacing Problem	28/01/2022	28/01/2022
								F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	28/01/2022	28/01/2022
										A071102	Failure to Interrogate	28/01/2022	28/01/2022
										A0712	Pacing Problem	28/01/2022	28/01/2022
								F1203	Life Threatening Illness or Injury	A070504	Premature Discharge of Battery	28/01/2022	28/01/2022
										A071102	Failure to Interrogate	28/01/2022	28/01/2022
										A0712	Pacing Problem	28/01/2022	28/01/2022
						CT321	Long-term surgical invasive	E011903	Syncope/Fainting	A070504	Premature Discharge of Battery	28/01/2022	28/01/2022
										A071102	Failure to Interrogate	28/01/2022	28/01/2022
										A0712	Pacing Problem	28/01/2022	28/01/2022
								E060106	Heart Block	A070504	Premature Discharge of Battery	28/01/2022	28/01/2022
										A071102	Failure to Interrogate	28/01/2022	28/01/2022
										A0712	Pacing Problem	28/01/2022	28/01/2022
								E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	28/01/2022	28/01/2022
										A071102	Failure to Interrogate	28/01/2022	28/01/2022
										A0712	Pacing Problem	28/01/2022	28/01/2022
								F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	28/01/2022	28/01/2022
										A071102	Failure to Interrogate	28/01/2022	28/01/2022
										A0712	Pacing Problem	28/01/2022	28/01/2022
								F1203	Life Threatening Illness or Injury	A070504	Premature Discharge of Battery	28/01/2022	28/01/2022
										A071102	Failure to Interrogate	28/01/2022	28/01/2022
										A0712	Pacing Problem	28/01/2022	28/01/2022
						CT449	Generators and associated devices	E011903	Syncope/Fainting	A070504	Premature Discharge of Battery	28/01/2022	28/01/2022

									A071102	Failure to Interrogate	28/01/2022	28/01/2022
									A0712	Pacing Problem	28/01/2022	28/01/2022
							E060106	Heart Block	A070504	Premature Discharge of Battery	28/01/2022	28/01/2022
									A071102	Failure to Interrogate	28/01/2022	28/01/2022
									A0712	Pacing Problem	28/01/2022	28/01/2022
							E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	28/01/2022	28/01/2022
									A071102	Failure to Interrogate	28/01/2022	28/01/2022
									A0712	Pacing Problem	28/01/2022	28/01/2022
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	28/01/2022	28/01/2022
									A071102	Failure to Interrogate	28/01/2022	28/01/2022
									A0712	Pacing Problem	28/01/2022	28/01/2022
							F1203	Life Threatening Illness or Injury	A070504	Premature Discharge of Battery	28/01/2022	28/01/2022
									A071102	Failure to Interrogate	28/01/2022	28/01/2022
									A0712	Pacing Problem	28/01/2022	28/01/2022
					CT981	Single use	E011903	Syncope/Fainting	A070504	Premature Discharge of Battery	28/01/2022	28/01/2022
									A071102	Failure to Interrogate	28/01/2022	28/01/2022
									A0712	Pacing Problem	28/01/2022	28/01/2022
							E060106	Heart Block	A070504	Premature Discharge of Battery	28/01/2022	28/01/2022
									A071102	Failure to Interrogate	28/01/2022	28/01/2022
									A0712	Pacing Problem	28/01/2022	28/01/2022
							E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	28/01/2022	28/01/2022
									A071102	Failure to Interrogate	28/01/2022	28/01/2022
									A0712	Pacing Problem	28/01/2022	28/01/2022
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	28/01/2022	28/01/2022
									A071102	Failure to Interrogate	28/01/2022	28/01/2022
									A0712	Pacing Problem	28/01/2022	28/01/2022
							F1203	Life Threatening Illness or Injury	A070504	Premature Discharge of Battery	28/01/2022	28/01/2022
									A071102	Failure to Interrogate	28/01/2022	28/01/2022
									A0712	Pacing Problem	28/01/2022	28/01/2022
					CT983	Surgical invasive	E011903	Syncope/Fainting	A070504	Premature Discharge of Battery	28/01/2022	28/01/2022
									A071102	Failure to Interrogate	28/01/2022	28/01/2022
									A0712	Pacing Problem	28/01/2022	28/01/2022
							E060106	Heart Block	A070504	Premature Discharge of Battery	28/01/2022	28/01/2022
									A071102	Failure to Interrogate	28/01/2022	28/01/2022
									A0712	Pacing Problem	28/01/2022	28/01/2022
							E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	28/01/2022	28/01/2022
									A071102	Failure to Interrogate	28/01/2022	28/01/2022
									A0712	Pacing Problem	28/01/2022	28/01/2022
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	28/01/2022	28/01/2022
									A071102	Failure to Interrogate	28/01/2022	28/01/2022
									A0712	Pacing Problem	28/01/2022	28/01/2022
							F1203	Life Threatening Illness or Injury	A070504	Premature Discharge of Battery	28/01/2022	28/01/2022
									A071102	Failure to Interrogate	28/01/2022	28/01/2022
									A0712	Pacing Problem	28/01/2022	28/01/2022
7732	2022	All other reportable incidents	37265	Dual-chamber implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A0722	Impedance Problem		
							F26	No Health Consequences or Impact	A0722	Impedance Problem		
7733	2022	All other reportable incidents	47263	Cardiac resynchronization therapy implantable pacemaker	CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	27/01/2022	27/01/2022
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	27/01/2022	27/01/2022
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	27/01/2022	27/01/2022
					CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	27/01/2022	27/01/2022
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	27/01/2022	27/01/2022
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	27/01/2022	27/01/2022
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	27/01/2022	27/01/2022
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	27/01/2022	27/01/2022
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	27/01/2022	27/01/2022
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	27/01/2022	27/01/2022
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	27/01/2022	27/01/2022
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	27/01/2022	27/01/2022
7734	2022	All other reportable incidents	47270	Cardiac resynchronization therapy implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A070909	Oversensing		
									A071302	Inappropriate/Inadequate Shock/Stimulation		
							F22	Unexpected Diagnostic Intervention	A070909	Oversensing		
									A071302	Inappropriate/Inadequate Shock/Stimulation		
7735	2022	All other reportable incidents	47265	Dual-chamber implantable pacemaker, rate-responsive	CT141	Electrodes/Leads and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	08/09/2022	08/09/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	08/09/2022	08/09/2022
					CT145	Cardiovascular devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	08/09/2022	08/09/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	08/09/2022	08/09/2022
					CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	08/09/2022	08/09/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	08/09/2022	08/09/2022
					CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	08/09/2022	08/09/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	08/09/2022	08/09/2022
					CT233	Surgical	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	08/09/2022	08/09/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	08/09/2022	08/09/2022
					CT2406	Implantable	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	08/09/2022	08/09/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	08/09/2022	08/09/2022
					CT2407	Active implantable	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	08/09/2022	08/09/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	08/09/2022	08/09/2022
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	08/09/2022	08/09/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	08/09/2022	08/09/2022
					CT2727	Active implant devices/ancillaries	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	08/09/2022	08/09/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	08/09/2022	08/09/2022
					CT2750	Implantable cardiac electrical stimulators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	08/09/2022	08/09/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	08/09/2022	08/09/2022
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	08/09/2022	08/09/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	08/09/2022	08/09/2022
					CT2985	Active	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	08/09/2022	08/09/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	08/09/2022	08/09/2022
					CT3001	Implantable metal	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	08/09/2022	08/09/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	08/09/2022	08/09/2022

					CT3002	Implantable synthetic polymer	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	08/09/2022	08/09/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	08/09/2022	08/09/2022
					CT3006	Implantable inorganic material	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	08/09/2022	08/09/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	08/09/2022	08/09/2022
					CT314	Electrically powered	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	08/09/2022	08/09/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	08/09/2022	08/09/2022
					CT321	Long-term surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	08/09/2022	08/09/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	08/09/2022	08/09/2022
					CT449	Generators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	08/09/2022	08/09/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	08/09/2022	08/09/2022
					CT981	Single use	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	08/09/2022	08/09/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	08/09/2022	08/09/2022
					CT983	Surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	08/09/2022	08/09/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	08/09/2022	08/09/2022
7736	2022	All other reportable incidents	17882	Physiologic-monitoring defibrillation system	CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070803	Failure to Power Up		
							F26	No Health Consequences or Impact	A070803	Failure to Power Up		
7737	2022	All other reportable incidents	36241	Pacemaker/defibrillator lead	CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A040101	Fracture		
									A072201	High Impedance		
							F2203	Imaging Required	A040101	Fracture		
									A072201	High Impedance		
7738	2022	All other reportable incidents	35852	Single-chamber implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	21/11/2022	21/11/2022
									A0908	Incorrect, Inadequate or Imprecise Result or Readings	21/11/2022	21/11/2022
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	21/11/2022	21/11/2022
									A0908	Incorrect, Inadequate or Imprecise Result or Readings	21/11/2022	21/11/2022
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	21/11/2022	21/11/2022
									A0908	Incorrect, Inadequate or Imprecise Result or Readings	21/11/2022	21/11/2022
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery	21/11/2022	21/11/2022
									A0908	Incorrect, Inadequate or Imprecise Result or Readings	21/11/2022	21/11/2022
7739	2022	All other reportable incidents	47267	Single-chamber implantable pacemaker, rate-responsive	CT141	Electrodes/Leads and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A0908	Incorrect, Inadequate or Imprecise Result or Readings	09/09/2022	09/09/2022
							F08	Hospitalization or Prolonged Hospitalization	A0908	Incorrect, Inadequate or Imprecise Result or Readings	09/09/2022	09/09/2022
							F1905	Device Revision or Replacement	A0908	Incorrect, Inadequate or Imprecise Result or Readings	09/09/2022	09/09/2022
					CT145	Cardiovascular devices	E2403	No Clinical Signs, Symptoms or Conditions	A0908	Incorrect, Inadequate or Imprecise Result or Readings	09/09/2022	09/09/2022
							F08	Hospitalization or Prolonged Hospitalization	A0908	Incorrect, Inadequate or Imprecise Result or Readings	09/09/2022	09/09/2022
							F1905	Device Revision or Replacement	A0908	Incorrect, Inadequate or Imprecise Result or Readings	09/09/2022	09/09/2022
					CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A0908	Incorrect, Inadequate or Imprecise Result or Readings	09/09/2022	09/09/2022
							F08	Hospitalization or Prolonged Hospitalization	A0908	Incorrect, Inadequate or Imprecise Result or Readings	09/09/2022	09/09/2022
							F1905	Device Revision or Replacement	A0908	Incorrect, Inadequate or Imprecise Result or Readings	09/09/2022	09/09/2022
					CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A0908	Incorrect, Inadequate or Imprecise Result or Readings	09/09/2022	09/09/2022
							F08	Hospitalization or Prolonged Hospitalization	A0908	Incorrect, Inadequate or Imprecise Result or Readings	09/09/2022	09/09/2022
							F1905	Device Revision or Replacement	A0908	Incorrect, Inadequate or Imprecise Result or Readings	09/09/2022	09/09/2022
					CT233	Surgical	E2403	No Clinical Signs, Symptoms or Conditions	A0908	Incorrect, Inadequate or Imprecise Result or Readings	09/09/2022	09/09/2022
							F08	Hospitalization or Prolonged Hospitalization	A0908	Incorrect, Inadequate or Imprecise Result or Readings	09/09/2022	09/09/2022
							F1905	Device Revision or Replacement	A0908	Incorrect, Inadequate or Imprecise Result or Readings	09/09/2022	09/09/2022
					CT2406	Implantable	E2403	No Clinical Signs, Symptoms or Conditions	A0908	Incorrect, Inadequate or Imprecise Result or Readings	09/09/2022	09/09/2022
							F08	Hospitalization or Prolonged Hospitalization	A0908	Incorrect, Inadequate or Imprecise Result or Readings	09/09/2022	09/09/2022
							F1905	Device Revision or Replacement	A0908	Incorrect, Inadequate or Imprecise Result or Readings	09/09/2022	09/09/2022
					CT2407	Active implantable	E2403	No Clinical Signs, Symptoms or Conditions	A0908	Incorrect, Inadequate or Imprecise Result or Readings	09/09/2022	09/09/2022
							F08	Hospitalization or Prolonged Hospitalization	A0908	Incorrect, Inadequate or Imprecise Result or Readings	09/09/2022	09/09/2022
							F1905	Device Revision or Replacement	A0908	Incorrect, Inadequate or Imprecise Result or Readings	09/09/2022	09/09/2022
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A0908	Incorrect, Inadequate or Imprecise Result or Readings	09/09/2022	09/09/2022
							F08	Hospitalization or Prolonged Hospitalization	A0908	Incorrect, Inadequate or Imprecise Result or Readings	09/09/2022	09/09/2022
							F1905	Device Revision or Replacement	A0908	Incorrect, Inadequate or Imprecise Result or Readings	09/09/2022	09/09/2022
					CT2727	Active implant devices/ancillaries	E2403	No Clinical Signs, Symptoms or Conditions	A0908	Incorrect, Inadequate or Imprecise Result or Readings	09/09/2022	09/09/2022
							F08	Hospitalization or Prolonged Hospitalization	A0908	Incorrect, Inadequate or Imprecise Result or Readings	09/09/2022	09/09/2022
							F1905	Device Revision or Replacement	A0908	Incorrect, Inadequate or Imprecise Result or Readings	09/09/2022	09/09/2022
					CT2750	Implantable cardiac electrical stimulators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A0908	Incorrect, Inadequate or Imprecise Result or Readings	09/09/2022	09/09/2022
							F08	Hospitalization or Prolonged Hospitalization	A0908	Incorrect, Inadequate or Imprecise Result or Readings	09/09/2022	09/09/2022
							F1905	Device Revision or Replacement	A0908	Incorrect, Inadequate or Imprecise Result or Readings	09/09/2022	09/09/2022
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A0908	Incorrect, Inadequate or Imprecise Result or Readings	09/09/2022	09/09/2022
							F08	Hospitalization or Prolonged Hospitalization	A0908	Incorrect, Inadequate or Imprecise Result or Readings	09/09/2022	09/09/2022

							F1905	Device Revision or Replacement	A0908	Incorrect, Inadequate or Imprecise Result or Readings	09/09/2022	09/09/2022
					CT2985	Active	E2403	No Clinical Signs, Symptoms or Conditions	A0908	Incorrect, Inadequate or Imprecise Result or Readings	09/09/2022	09/09/2022
							F08	Hospitalization or Prolonged Hospitalization	A0908	Incorrect, Inadequate or Imprecise Result or Readings	09/09/2022	09/09/2022
							F1905	Device Revision or Replacement	A0908	Incorrect, Inadequate or Imprecise Result or Readings	09/09/2022	09/09/2022
					CT3001	Implantable metal	E2403	No Clinical Signs, Symptoms or Conditions	A0908	Incorrect, Inadequate or Imprecise Result or Readings	09/09/2022	09/09/2022
							F08	Hospitalization or Prolonged Hospitalization	A0908	Incorrect, Inadequate or Imprecise Result or Readings	09/09/2022	09/09/2022
							F1905	Device Revision or Replacement	A0908	Incorrect, Inadequate or Imprecise Result or Readings	09/09/2022	09/09/2022
					CT3002	Implantable synthetic polymer	E2403	No Clinical Signs, Symptoms or Conditions	A0908	Incorrect, Inadequate or Imprecise Result or Readings	09/09/2022	09/09/2022
							F08	Hospitalization or Prolonged Hospitalization	A0908	Incorrect, Inadequate or Imprecise Result or Readings	09/09/2022	09/09/2022
							F1905	Device Revision or Replacement	A0908	Incorrect, Inadequate or Imprecise Result or Readings	09/09/2022	09/09/2022
					CT3006	Implantable inorganic material	E2403	No Clinical Signs, Symptoms or Conditions	A0908	Incorrect, Inadequate or Imprecise Result or Readings	09/09/2022	09/09/2022
							F08	Hospitalization or Prolonged Hospitalization	A0908	Incorrect, Inadequate or Imprecise Result or Readings	09/09/2022	09/09/2022
							F1905	Device Revision or Replacement	A0908	Incorrect, Inadequate or Imprecise Result or Readings	09/09/2022	09/09/2022
					CT314	Electrically powered	E2403	No Clinical Signs, Symptoms or Conditions	A0908	Incorrect, Inadequate or Imprecise Result or Readings	09/09/2022	09/09/2022
							F08	Hospitalization or Prolonged Hospitalization	A0908	Incorrect, Inadequate or Imprecise Result or Readings	09/09/2022	09/09/2022
							F1905	Device Revision or Replacement	A0908	Incorrect, Inadequate or Imprecise Result or Readings	09/09/2022	09/09/2022
					CT321	Long-term surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A0908	Incorrect, Inadequate or Imprecise Result or Readings	09/09/2022	09/09/2022
							F08	Hospitalization or Prolonged Hospitalization	A0908	Incorrect, Inadequate or Imprecise Result or Readings	09/09/2022	09/09/2022
							F1905	Device Revision or Replacement	A0908	Incorrect, Inadequate or Imprecise Result or Readings	09/09/2022	09/09/2022
					CT449	Generators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A0908	Incorrect, Inadequate or Imprecise Result or Readings	09/09/2022	09/09/2022
							F08	Hospitalization or Prolonged Hospitalization	A0908	Incorrect, Inadequate or Imprecise Result or Readings	09/09/2022	09/09/2022
							F1905	Device Revision or Replacement	A0908	Incorrect, Inadequate or Imprecise Result or Readings	09/09/2022	09/09/2022
					CT981	Single use	E2403	No Clinical Signs, Symptoms or Conditions	A0908	Incorrect, Inadequate or Imprecise Result or Readings	09/09/2022	09/09/2022
							F08	Hospitalization or Prolonged Hospitalization	A0908	Incorrect, Inadequate or Imprecise Result or Readings	09/09/2022	09/09/2022
							F1905	Device Revision or Replacement	A0908	Incorrect, Inadequate or Imprecise Result or Readings	09/09/2022	09/09/2022
					CT983	Surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A0908	Incorrect, Inadequate or Imprecise Result or Readings	09/09/2022	09/09/2022
							F08	Hospitalization or Prolonged Hospitalization	A0908	Incorrect, Inadequate or Imprecise Result or Readings	09/09/2022	09/09/2022
							F1905	Device Revision or Replacement	A0908	Incorrect, Inadequate or Imprecise Result or Readings	09/09/2022	09/09/2022
7740	2022	All other reportable incidents	35852	Single-chamber implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	27/10/2022	27/10/2022
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	27/10/2022	27/10/2022
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	27/10/2022	27/10/2022
							F24	Insufficient Health Impact Information	A070504	Premature Discharge of Battery	27/10/2022	27/10/2022
7741	2022	All other reportable incidents	47270	Cardiac resynchronization therapy implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A160503	Inappropriate or Unexpected Reset	26/01/2022	26/01/2022
									A17	Compatibility Problem	26/01/2022	26/01/2022
							F23	Unexpected Medical Intervention	A160503	Inappropriate or Unexpected Reset	26/01/2022	26/01/2022
									A17	Compatibility Problem	26/01/2022	26/01/2022
7742	2022	All other reportable incidents	35223	Endocardial/interventricular septal pacing lead	CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A0401	Break	02/09/2022	02/09/2022
							F1908	Prolonged Surgery	A0401	Break	02/09/2022	02/09/2022
7743	2022	All other reportable incidents	47265	Dual-chamber implantable pacemaker, rate-responsive	CT141	Electrodes/Leads and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	09/09/2022	09/09/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	09/09/2022	09/09/2022
					CT145	Cardiovascular devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	09/09/2022	09/09/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	09/09/2022	09/09/2022
					CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	09/09/2022	09/09/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	09/09/2022	09/09/2022
					CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	09/09/2022	09/09/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	09/09/2022	09/09/2022
					CT233	Surgical	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	09/09/2022	09/09/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	09/09/2022	09/09/2022
					CT2406	Implantable	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	09/09/2022	09/09/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	09/09/2022	09/09/2022
					CT2407	Active implantable	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	09/09/2022	09/09/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	09/09/2022	09/09/2022
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	09/09/2022	09/09/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	09/09/2022	09/09/2022
					CT2727	Active implant devices/ancillaries	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	09/09/2022	09/09/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	09/09/2022	09/09/2022
					CT2750	Implantable cardiac electrical stimulators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	09/09/2022	09/09/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	09/09/2022	09/09/2022
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	09/09/2022	09/09/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	09/09/2022	09/09/2022
					CT2985	Active	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	09/09/2022	09/09/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	09/09/2022	09/09/2022
					CT3001	Implantable metal	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	09/09/2022	09/09/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	09/09/2022	09/09/2022
					CT3002	Implantable synthetic polymer	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	09/09/2022	09/09/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	09/09/2022	09/09/2022

					CT3006	Implantable inorganic material	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	09/09/2022	09/09/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	09/09/2022	09/09/2022
					CT314	Electrically powered	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	09/09/2022	09/09/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	09/09/2022	09/09/2022
					CT321	Long-term surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	09/09/2022	09/09/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	09/09/2022	09/09/2022
					CT449	Generators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	09/09/2022	09/09/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	09/09/2022	09/09/2022
					CT981	Single use	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	09/09/2022	09/09/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	09/09/2022	09/09/2022
					CT983	Surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	09/09/2022	09/09/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	09/09/2022	09/09/2022
7744	2022	All other reportable incidents	60789	Intracardiac pacemaker	CT130	Catheters and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070102	High Capture Threshold		
									A070504	Premature Discharge of Battery		
									A072202	Low Impedance		
							F08	Hospitalization or Prolonged Hospitalization	A070102	High Capture Threshold		
									A070504	Premature Discharge of Battery		
									A072202	Low Impedance		
							F12	Serious Injury/ Illness/ Impairment	A070102	High Capture Threshold		
									A070504	Premature Discharge of Battery		
									A072202	Low Impedance		
							F1905	Device Revision or Replacement	A070102	High Capture Threshold		
									A070504	Premature Discharge of Battery		
									A072202	Low Impedance		
							F26	No Health Consequences or Impact	A070102	High Capture Threshold		
									A070504	Premature Discharge of Battery		
									A072202	Low Impedance		
					CT145	Cardiovascular devices	E2403	No Clinical Signs, Symptoms or Conditions	A070102	High Capture Threshold		
									A070504	Premature Discharge of Battery		
									A072202	Low Impedance		
							F08	Hospitalization or Prolonged Hospitalization	A070102	High Capture Threshold		
									A070504	Premature Discharge of Battery		
									A072202	Low Impedance		
							F12	Serious Injury/ Illness/ Impairment	A070102	High Capture Threshold		
									A070504	Premature Discharge of Battery		
									A072202	Low Impedance		
							F1905	Device Revision or Replacement	A070102	High Capture Threshold		
									A070504	Premature Discharge of Battery		
									A072202	Low Impedance		
							F26	No Health Consequences or Impact	A070102	High Capture Threshold		
									A070504	Premature Discharge of Battery		
									A072202	Low Impedance		
					CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070102	High Capture Threshold		
									A070504	Premature Discharge of Battery		
									A072202	Low Impedance		
							F08	Hospitalization or Prolonged Hospitalization	A070102	High Capture Threshold		
									A070504	Premature Discharge of Battery		
									A072202	Low Impedance		
							F12	Serious Injury/ Illness/ Impairment	A070102	High Capture Threshold		
									A070504	Premature Discharge of Battery		
									A072202	Low Impedance		
							F1905	Device Revision or Replacement	A070102	High Capture Threshold		
									A070504	Premature Discharge of Battery		
									A072202	Low Impedance		
							F26	No Health Consequences or Impact	A070102	High Capture Threshold		
									A070504	Premature Discharge of Battery		
									A072202	Low Impedance		
					CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A070102	High Capture Threshold		
									A070504	Premature Discharge of Battery		
									A072202	Low Impedance		
							F08	Hospitalization or Prolonged Hospitalization	A070102	High Capture Threshold		
									A070504	Premature Discharge of Battery		
									A072202	Low Impedance		
							F12	Serious Injury/ Illness/ Impairment	A070102	High Capture Threshold		
									A070504	Premature Discharge of Battery		
									A072202	Low Impedance		
							F1905	Device Revision or Replacement	A070102	High Capture Threshold		
									A070504	Premature Discharge of Battery		
									A072202	Low Impedance		
							F26	No Health Consequences or Impact	A070102	High Capture Threshold		
									A070504	Premature Discharge of Battery		
									A072202	Low Impedance		
					CT233	Surgical	E2403	No Clinical Signs, Symptoms or Conditions	A070102	High Capture Threshold		
									A070504	Premature Discharge of Battery		
									A072202	Low Impedance		
							F08	Hospitalization or Prolonged Hospitalization	A070102	High Capture Threshold		
									A070504	Premature Discharge of Battery		
									A072202	Low Impedance		
							F12	Serious Injury/ Illness/ Impairment	A070102	High Capture Threshold		
									A070504	Premature Discharge of Battery		
									A072202	Low Impedance		
							F1905	Device Revision or Replacement	A070102	High Capture Threshold		
									A070504	Premature Discharge of Battery		
									A072202	Low Impedance		
							F26	No Health Consequences or Impact	A070102	High Capture Threshold		
									A070504	Premature Discharge of Battery		
									A072202	Low Impedance		
					CT2406	Implantable	E2403	No Clinical Signs, Symptoms or Conditions	A070102	High Capture Threshold		

									A070504	Premature Discharge of Battery		
									A072202	Low Impedance		
							F08	Hospitalization or Prolonged Hospitalization	A070102	High Capture Threshold		
									A070504	Premature Discharge of Battery		
									A072202	Low Impedance		
							F12	Serious Injury/ Illness/ Impairment	A070102	High Capture Threshold		
									A070504	Premature Discharge of Battery		
									A072202	Low Impedance		
							F1905	Device Revision or Replacement	A070102	High Capture Threshold		
									A070504	Premature Discharge of Battery		
									A072202	Low Impedance		
							F26	No Health Consequences or Impact	A070102	High Capture Threshold		
									A070504	Premature Discharge of Battery		
									A072202	Low Impedance		
						CT2407	Active implantable	E2403	No Clinical Signs, Symptoms or Conditions	A070102	High Capture Threshold		
										A070504	Premature Discharge of Battery		
										A072202	Low Impedance		
								F08	Hospitalization or Prolonged Hospitalization	A070102	High Capture Threshold		
										A070504	Premature Discharge of Battery		
										A072202	Low Impedance		
								F12	Serious Injury/ Illness/ Impairment	A070102	High Capture Threshold		
										A070504	Premature Discharge of Battery		
										A072202	Low Impedance		
								F1905	Device Revision or Replacement	A070102	High Capture Threshold		
										A070504	Premature Discharge of Battery		
										A072202	Low Impedance		
								F26	No Health Consequences or Impact	A070102	High Capture Threshold		
										A070504	Premature Discharge of Battery		
										A072202	Low Impedance		
						CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A070102	High Capture Threshold		
										A070504	Premature Discharge of Battery		
										A072202	Low Impedance		
								F08	Hospitalization or Prolonged Hospitalization	A070102	High Capture Threshold		
										A070504	Premature Discharge of Battery		
										A072202	Low Impedance		
								F12	Serious Injury/ Illness/ Impairment	A070102	High Capture Threshold		
										A070504	Premature Discharge of Battery		
										A072202	Low Impedance		
								F1905	Device Revision or Replacement	A070102	High Capture Threshold		
										A070504	Premature Discharge of Battery		
										A072202	Low Impedance		
								F26	No Health Consequences or Impact	A070102	High Capture Threshold		
										A070504	Premature Discharge of Battery		
										A072202	Low Impedance		
						CT2727	Active implant devices/ancillaries	E2403	No Clinical Signs, Symptoms or Conditions	A070102	High Capture Threshold		
										A070504	Premature Discharge of Battery		
										A072202	Low Impedance		
								F08	Hospitalization or Prolonged Hospitalization	A070102	High Capture Threshold		
										A070504	Premature Discharge of Battery		
										A072202	Low Impedance		
								F12	Serious Injury/ Illness/ Impairment	A070102	High Capture Threshold		
										A070504	Premature Discharge of Battery		
										A072202	Low Impedance		
								F1905	Device Revision or Replacement	A070102	High Capture Threshold		
										A070504	Premature Discharge of Battery		
										A072202	Low Impedance		
								F26	No Health Consequences or Impact	A070102	High Capture Threshold		
										A070504	Premature Discharge of Battery		
										A072202	Low Impedance		
						CT2750	Implantable cardiac electrical stimulators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070102	High Capture Threshold		
										A070504	Premature Discharge of Battery		
										A072202	Low Impedance		
								F08	Hospitalization or Prolonged Hospitalization	A070102	High Capture Threshold		
										A070504	Premature Discharge of Battery		
										A072202	Low Impedance		
								F12	Serious Injury/ Illness/ Impairment	A070102	High Capture Threshold		
										A070504	Premature Discharge of Battery		
										A072202	Low Impedance		
								F1905	Device Revision or Replacement	A070102	High Capture Threshold		
										A070504	Premature Discharge of Battery		
										A072202	Low Impedance		
								F26	No Health Consequences or Impact	A070102	High Capture Threshold		
										A070504	Premature Discharge of Battery		
										A072202	Low Impedance		
						CT2818	Catheterization/Cannulation/Fluid path devices	E2403	No Clinical Signs, Symptoms or Conditions	A070102	High Capture Threshold		
										A070504	Premature Discharge of Battery		
										A072202	Low Impedance		
								F08	Hospitalization or Prolonged Hospitalization	A070102	High Capture Threshold		
										A070504	Premature Discharge of Battery		
										A072202	Low Impedance		
								F12	Serious Injury/ Illness/ Impairment	A070102	High Capture Threshold		
										A070504	Premature Discharge of Battery		
										A072202	Low Impedance		
								F1905	Device Revision or Replacement	A070102	High Capture Threshold		
										A070504	Premature Discharge of Battery		
										A072202	Low Impedance		
								F26	No Health Consequences or Impact	A070102	High Capture Threshold		
										A070504	Premature Discharge of Battery		

								A072202	Low Impedance		
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A070102	High Capture Threshold		
									A070504	Premature Discharge of Battery		
									A072202	Low Impedance		
							F08	Hospitalization or Prolonged Hospitalization	A070102	High Capture Threshold		
									A070504	Premature Discharge of Battery		
									A072202	Low Impedance		
							F12	Serious Injury/ Illness/ Impairment	A070102	High Capture Threshold		
									A070504	Premature Discharge of Battery		
									A072202	Low Impedance		
							F1905	Device Revision or Replacement	A070102	High Capture Threshold		
									A070504	Premature Discharge of Battery		
									A072202	Low Impedance		
							F26	No Health Consequences or Impact	A070102	High Capture Threshold		
									A070504	Premature Discharge of Battery		
									A072202	Low Impedance		
					CT2985	Active	E2403	No Clinical Signs, Symptoms or Conditions	A070102	High Capture Threshold		
									A070504	Premature Discharge of Battery		
									A072202	Low Impedance		
							F08	Hospitalization or Prolonged Hospitalization	A070102	High Capture Threshold		
									A070504	Premature Discharge of Battery		
									A072202	Low Impedance		
							F12	Serious Injury/ Illness/ Impairment	A070102	High Capture Threshold		
									A070504	Premature Discharge of Battery		
									A072202	Low Impedance		
							F1905	Device Revision or Replacement	A070102	High Capture Threshold		
									A070504	Premature Discharge of Battery		
									A072202	Low Impedance		
							F26	No Health Consequences or Impact	A070102	High Capture Threshold		
									A070504	Premature Discharge of Battery		
									A072202	Low Impedance		
					CT3001	Implantable metal	E2403	No Clinical Signs, Symptoms or Conditions	A070102	High Capture Threshold		
									A070504	Premature Discharge of Battery		
									A072202	Low Impedance		
							F08	Hospitalization or Prolonged Hospitalization	A070102	High Capture Threshold		
									A070504	Premature Discharge of Battery		
									A072202	Low Impedance		
							F12	Serious Injury/ Illness/ Impairment	A070102	High Capture Threshold		
									A070504	Premature Discharge of Battery		
									A072202	Low Impedance		
							F1905	Device Revision or Replacement	A070102	High Capture Threshold		
									A070504	Premature Discharge of Battery		
									A072202	Low Impedance		
							F26	No Health Consequences or Impact	A070102	High Capture Threshold		
									A070504	Premature Discharge of Battery		
									A072202	Low Impedance		
					CT3002	Implantable synthetic polymer	E2403	No Clinical Signs, Symptoms or Conditions	A070102	High Capture Threshold		
									A070504	Premature Discharge of Battery		
									A072202	Low Impedance		
							F08	Hospitalization or Prolonged Hospitalization	A070102	High Capture Threshold		
									A070504	Premature Discharge of Battery		
									A072202	Low Impedance		
							F12	Serious Injury/ Illness/ Impairment	A070102	High Capture Threshold		
									A070504	Premature Discharge of Battery		
									A072202	Low Impedance		
							F1905	Device Revision or Replacement	A070102	High Capture Threshold		
									A070504	Premature Discharge of Battery		
									A072202	Low Impedance		
							F26	No Health Consequences or Impact	A070102	High Capture Threshold		
									A070504	Premature Discharge of Battery		
									A072202	Low Impedance		
					CT3006	Implantable inorganic material	E2403	No Clinical Signs, Symptoms or Conditions	A070102	High Capture Threshold		
									A070504	Premature Discharge of Battery		
									A072202	Low Impedance		
							F08	Hospitalization or Prolonged Hospitalization	A070102	High Capture Threshold		
									A070504	Premature Discharge of Battery		
									A072202	Low Impedance		
							F12	Serious Injury/ Illness/ Impairment	A070102	High Capture Threshold		
									A070504	Premature Discharge of Battery		
									A072202	Low Impedance		
							F1905	Device Revision or Replacement	A070102	High Capture Threshold		
									A070504	Premature Discharge of Battery		
									A072202	Low Impedance		
							F26	No Health Consequences or Impact	A070102	High Capture Threshold		
									A070504	Premature Discharge of Battery		
									A072202	Low Impedance		
					CT314	Electrically powered	E2403	No Clinical Signs, Symptoms or Conditions	A070102	High Capture Threshold		
									A070504	Premature Discharge of Battery		
									A072202	Low Impedance		
							F08	Hospitalization or Prolonged Hospitalization	A070102	High Capture Threshold		
									A070504	Premature Discharge of Battery		
									A072202	Low Impedance		
							F12	Serious Injury/ Illness/ Impairment	A070102	High Capture Threshold		
									A070504	Premature Discharge of Battery		
									A072202	Low Impedance		
							F1905	Device Revision or Replacement	A070102	High Capture Threshold		
									A070504	Premature Discharge of Battery		
									A072202	Low Impedance		

							F26	No Health Consequences or Impact	A070102	High Capture Threshold		
									A070504	Premature Discharge of Battery		
									A072202	Low Impedance		
					CT321	Long-term surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A070102	High Capture Threshold		
									A070504	Premature Discharge of Battery		
									A072202	Low Impedance		
							F08	Hospitalization or Prolonged Hospitalization	A070102	High Capture Threshold		
									A070504	Premature Discharge of Battery		
									A072202	Low Impedance		
							F12	Serious Injury/ Illness/ Impairment	A070102	High Capture Threshold		
									A070504	Premature Discharge of Battery		
									A072202	Low Impedance		
							F1905	Device Revision or Replacement	A070102	High Capture Threshold		
									A070504	Premature Discharge of Battery		
									A072202	Low Impedance		
							F26	No Health Consequences or Impact	A070102	High Capture Threshold		
									A070504	Premature Discharge of Battery		
									A072202	Low Impedance		
					CT981	Single use	E2403	No Clinical Signs, Symptoms or Conditions	A070102	High Capture Threshold		
									A070504	Premature Discharge of Battery		
									A072202	Low Impedance		
							F08	Hospitalization or Prolonged Hospitalization	A070102	High Capture Threshold		
									A070504	Premature Discharge of Battery		
									A072202	Low Impedance		
							F12	Serious Injury/ Illness/ Impairment	A070102	High Capture Threshold		
									A070504	Premature Discharge of Battery		
									A072202	Low Impedance		
							F1905	Device Revision or Replacement	A070102	High Capture Threshold		
									A070504	Premature Discharge of Battery		
									A072202	Low Impedance		
							F26	No Health Consequences or Impact	A070102	High Capture Threshold		
									A070504	Premature Discharge of Battery		
									A072202	Low Impedance		
					CT983	Surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A070102	High Capture Threshold		
									A070504	Premature Discharge of Battery		
									A072202	Low Impedance		
							F08	Hospitalization or Prolonged Hospitalization	A070102	High Capture Threshold		
									A070504	Premature Discharge of Battery		
									A072202	Low Impedance		
							F12	Serious Injury/ Illness/ Impairment	A070102	High Capture Threshold		
									A070504	Premature Discharge of Battery		
									A072202	Low Impedance		
							F1905	Device Revision or Replacement	A070102	High Capture Threshold		
									A070504	Premature Discharge of Battery		
									A072202	Low Impedance		
							F26	No Health Consequences or Impact	A070102	High Capture Threshold		
									A070504	Premature Discharge of Battery		
									A072202	Low Impedance		
7745	2022	All other reportable incidents			CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	21/09/2022	21/09/2022
									A0908	Incorrect, Inadequate or Imprecise Result or Readings	21/09/2022	21/09/2022
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	21/09/2022	21/09/2022
									A0908	Incorrect, Inadequate or Imprecise Result or Readings	21/09/2022	21/09/2022
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	21/09/2022	21/09/2022
									A0908	Incorrect, Inadequate or Imprecise Result or Readings	21/09/2022	21/09/2022
7746	2022	All other reportable incidents	47265	Dual-chamber implantable pacemaker, rate-responsive	CT141	Electrodes/Leads and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
					CT145	Cardiovascular devices	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
					CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
					CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
					CT233	Surgical	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
					CT2406	Implantable	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
					CT2407	Active implantable	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A0705	Battery Problem		

								A070504	Premature Discharge of Battery			
				CT2727	Active implant devices/ancillaries	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem			
								A070504	Premature Discharge of Battery			
						F26	No Health Consequences or Impact	A0705	Battery Problem			
								A070504	Premature Discharge of Battery			
				CT2750	Implantable cardiac electrical stimulators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem			
								A070504	Premature Discharge of Battery			
						F26	No Health Consequences or Impact	A0705	Battery Problem			
								A070504	Premature Discharge of Battery			
				CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem			
								A070504	Premature Discharge of Battery			
						F26	No Health Consequences or Impact	A0705	Battery Problem			
								A070504	Premature Discharge of Battery			
				CT2985	Active	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem			
								A070504	Premature Discharge of Battery			
						F26	No Health Consequences or Impact	A0705	Battery Problem			
								A070504	Premature Discharge of Battery			
				CT3001	Implantable metal	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem			
								A070504	Premature Discharge of Battery			
						F26	No Health Consequences or Impact	A0705	Battery Problem			
								A070504	Premature Discharge of Battery			
				CT3002	Implantable synthetic polymer	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem			
								A070504	Premature Discharge of Battery			
						F26	No Health Consequences or Impact	A0705	Battery Problem			
								A070504	Premature Discharge of Battery			
				CT3006	Implantable inorganic material	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem			
								A070504	Premature Discharge of Battery			
						F26	No Health Consequences or Impact	A0705	Battery Problem			
								A070504	Premature Discharge of Battery			
				CT314	Electrically powered	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem			
								A070504	Premature Discharge of Battery			
						F26	No Health Consequences or Impact	A0705	Battery Problem			
								A070504	Premature Discharge of Battery			
				CT321	Long-term surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem			
								A070504	Premature Discharge of Battery			
						F26	No Health Consequences or Impact	A0705	Battery Problem			
								A070504	Premature Discharge of Battery			
				CT449	Generators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem			
								A070504	Premature Discharge of Battery			
						F26	No Health Consequences or Impact	A0705	Battery Problem			
								A070504	Premature Discharge of Battery			
				CT981	Single use	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem			
								A070504	Premature Discharge of Battery			
						F26	No Health Consequences or Impact	A0705	Battery Problem			
								A070504	Premature Discharge of Battery			
				CT983	Surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem			
								A070504	Premature Discharge of Battery			
						F26	No Health Consequences or Impact	A0705	Battery Problem			
								A070504	Premature Discharge of Battery			
7747	2022	All other reportable incidents	35852	Single-chamber implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem	02/02/2022	02/02/2022
									A070504	Premature Discharge of Battery	02/02/2022	02/02/2022
						F08	Hospitalization or Prolonged Hospitalization	A0705	Battery Problem	02/02/2022	02/02/2022	
								A070504	Premature Discharge of Battery	02/02/2022	02/02/2022	
						F1905	Device Revision or Replacement	A0705	Battery Problem	02/02/2022	02/02/2022	
								A070504	Premature Discharge of Battery	02/02/2022	02/02/2022	
7748	2022	All other reportable incidents	36241	Pacemaker/defibrillator lead	CT1684	Cardiac pacemakers and associated devices	E2008	Foreign Body In Patient	A040101	Fracture	13/08/2021	13/08/2021
									A150207	Difficult to Remove	13/08/2021	13/08/2021
							E2403	No Clinical Signs, Symptoms or Conditions	A040101	Fracture	13/08/2021	13/08/2021
									A150207	Difficult to Remove	13/08/2021	13/08/2021
							F15	Recognized Device or Procedural Complication	A040101	Fracture	13/08/2021	13/08/2021
									A150207	Difficult to Remove	13/08/2021	13/08/2021
							F1905	Device Revision or Replacement	A040101	Fracture	13/08/2021	13/08/2021
									A150207	Difficult to Remove	13/08/2021	13/08/2021
7749	2022	All other reportable incidents	47265	Dual-chamber implantable pacemaker, rate-responsive	CT141	Electrodes/Leads and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	06/09/2022	06/09/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	06/09/2022	06/09/2022
					CT145	Cardiovascular devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	06/09/2022	06/09/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	06/09/2022	06/09/2022
					CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	06/09/2022	06/09/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	06/09/2022	06/09/2022
					CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	06/09/2022	06/09/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	06/09/2022	06/09/2022
					CT233	Surgical	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	06/09/2022	06/09/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	06/09/2022	06/09/2022
					CT2406	Implantable	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	06/09/2022	06/09/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	06/09/2022	06/09/2022
					CT2407	Active implantable	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	06/09/2022	06/09/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	06/09/2022	06/09/2022
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	06/09/2022	06/09/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	06/09/2022	06/09/2022
					CT2727	Active implant devices/ancillaries	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	06/09/2022	06/09/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	06/09/2022	06/09/2022
					CT2750	Implantable cardiac electrical stimulators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	06/09/2022	06/09/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	06/09/2022	06/09/2022
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	06/09/2022	06/09/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	06/09/2022	06/09/2022
					CT2985	Active	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	06/09/2022	06/09/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	06/09/2022	06/09/2022
					CT3001	Implantable metal	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	06/09/2022	06/09/2022

						F1901	Additional Surgery	A24	No Apparent Device Problem	06/09/2022	06/09/2022
					CT3002	Implantable synthetic polymer	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	06/09/2022	06/09/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	06/09/2022	06/09/2022
					CT3006	Implantable inorganic material	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	06/09/2022	06/09/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	06/09/2022	06/09/2022
					CT314	Electrically powered	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	06/09/2022	06/09/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	06/09/2022	06/09/2022
					CT321	Long-term surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	06/09/2022	06/09/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	06/09/2022	06/09/2022
					CT449	Generators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	06/09/2022	06/09/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	06/09/2022	06/09/2022
					CT981	Single use	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	06/09/2022	06/09/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	06/09/2022	06/09/2022
					CT983	Surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	06/09/2022	06/09/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	06/09/2022	06/09/2022
7750	2022		47265	Dual-chamber implantable pacemaker, rate-responsive	CT141	Electrodes/Leads and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	10/09/2022	10/09/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	10/09/2022	10/09/2022
					CT145	Cardiovascular devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	10/09/2022	10/09/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	10/09/2022	10/09/2022
					CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	10/09/2022	10/09/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	10/09/2022	10/09/2022
					CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	10/09/2022	10/09/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	10/09/2022	10/09/2022
					CT233	Surgical	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	10/09/2022	10/09/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	10/09/2022	10/09/2022
					CT2406	Implantable	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	10/09/2022	10/09/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	10/09/2022	10/09/2022
					CT2407	Active implantable	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	10/09/2022	10/09/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	10/09/2022	10/09/2022
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	10/09/2022	10/09/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	10/09/2022	10/09/2022
					CT2727	Active implant devices/ancillaries	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	10/09/2022	10/09/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	10/09/2022	10/09/2022
					CT2750	Implantable cardiac electrical stimulators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	10/09/2022	10/09/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	10/09/2022	10/09/2022
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	10/09/2022	10/09/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	10/09/2022	10/09/2022
					CT2985	Active	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	10/09/2022	10/09/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	10/09/2022	10/09/2022
					CT3001	Implantable metal	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	10/09/2022	10/09/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	10/09/2022	10/09/2022
					CT3002	Implantable synthetic polymer	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	10/09/2022	10/09/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	10/09/2022	10/09/2022
					CT3006	Implantable inorganic material	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	10/09/2022	10/09/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	10/09/2022	10/09/2022
					CT314	Electrically powered	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	10/09/2022	10/09/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	10/09/2022	10/09/2022
					CT321	Long-term surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	10/09/2022	10/09/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	10/09/2022	10/09/2022
					CT449	Generators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	10/09/2022	10/09/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	10/09/2022	10/09/2022
					CT981	Single use	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	10/09/2022	10/09/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	10/09/2022	10/09/2022
					CT983	Surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	10/09/2022	10/09/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	10/09/2022	10/09/2022
7751	2022		47265	Dual-chamber implantable pacemaker, rate-responsive	CT141	Electrodes/Leads and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	10/09/2022	10/09/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	10/09/2022	10/09/2022
					CT145	Cardiovascular devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	10/09/2022	10/09/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	10/09/2022	10/09/2022
					CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	10/09/2022	10/09/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	10/09/2022	10/09/2022
					CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	10/09/2022	10/09/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	10/09/2022	10/09/2022
					CT233	Surgical	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	10/09/2022	10/09/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	10/09/2022	10/09/2022
					CT2406	Implantable	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	10/09/2022	10/09/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	10/09/2022	10/09/2022
					CT2407	Active implantable	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	10/09/2022	10/09/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	10/09/2022	10/09/2022
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	10/09/2022	10/09/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	10/09/2022	10/09/2022
					CT2727	Active implant devices/ancillaries	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	10/09/2022	10/09/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	10/09/2022	10/09/2022
					CT2750	Implantable cardiac electrical stimulators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	10/09/2022	10/09/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	10/09/2022	10/09/2022
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	10/09/2022	10/09/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	10/09/2022	10/09/2022
					CT2985	Active	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	10/09/2022	10/09/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	10/09/2022	10/09/2022
					CT3001	Implantable metal	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	10/09/2022	10/09/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	10/09/2022	10/09/2022
					CT3002	Implantable synthetic polymer	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	10/09/2022	10/09/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	10/09/2022	10/09/2022
					CT3006	Implantable inorganic material	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	10/09/2022	10/09/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	10/09/2022	10/09/2022
					CT314	Electrically powered	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	10/09/2022	10/09/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	10/09/2022	10/09/2022
					CT321	Long-term surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	10/09/2022	10/09/2022

							F1901	Additional Surgery	A24	No Apparent Device Problem	10/09/2022	10/09/2022
					CT449	Generators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	10/09/2022	10/09/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	10/09/2022	10/09/2022
					CT981	Single use	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	10/09/2022	10/09/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	10/09/2022	10/09/2022
					CT983	Surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	10/09/2022	10/09/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	10/09/2022	10/09/2022
7752	2022	All other reportable incidents	35852	Single-chamber implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E2104	Electric Shock	A0712	Pacing Problem		
									A071302	Inappropriate/Inadequate Shock/Stimulation		
							E2403	No Clinical Signs, Symptoms or Conditions	A0712	Pacing Problem		
									A071302	Inappropriate/Inadequate Shock/Stimulation		
							F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A0712	Pacing Problem		
									A071302	Inappropriate/Inadequate Shock/Stimulation		
7753	2022	All other reportable incidents	47265	Dual-chamber implantable pacemaker, rate-responsive	CT141	Electrodes/Leads and associated devices	E011903	Syncope/Fainting	A0104	Migration or Expulsion of Device		
							E060101	Asystole	A0104	Migration or Expulsion of Device		
							E2007	Fall	A0104	Migration or Expulsion of Device		
							E2403	No Clinical Signs, Symptoms or Conditions	A0104	Migration or Expulsion of Device		
							F08	Hospitalization or Prolonged Hospitalization	A0104	Migration or Expulsion of Device		
							F1203	Life Threatening Illness or Injury	A0104	Migration or Expulsion of Device		
							F23	Unexpected Medical Intervention	A0104	Migration or Expulsion of Device		
					CT145	Cardiovascular devices	E011903	Syncope/Fainting	A0104	Migration or Expulsion of Device		
							E060101	Asystole	A0104	Migration or Expulsion of Device		
							E2007	Fall	A0104	Migration or Expulsion of Device		
							E2403	No Clinical Signs, Symptoms or Conditions	A0104	Migration or Expulsion of Device		
							F08	Hospitalization or Prolonged Hospitalization	A0104	Migration or Expulsion of Device		
							F1203	Life Threatening Illness or Injury	A0104	Migration or Expulsion of Device		
							F23	Unexpected Medical Intervention	A0104	Migration or Expulsion of Device		
					CT1684	Cardiac pacemakers and associated devices	E011903	Syncope/Fainting	A0104	Migration or Expulsion of Device		
							E060101	Asystole	A0104	Migration or Expulsion of Device		
							E2007	Fall	A0104	Migration or Expulsion of Device		
							E2403	No Clinical Signs, Symptoms or Conditions	A0104	Migration or Expulsion of Device		
							F08	Hospitalization or Prolonged Hospitalization	A0104	Migration or Expulsion of Device		
							F1203	Life Threatening Illness or Injury	A0104	Migration or Expulsion of Device		
							F23	Unexpected Medical Intervention	A0104	Migration or Expulsion of Device		
					CT1685	Cardiac pacemakers	E011903	Syncope/Fainting	A0104	Migration or Expulsion of Device		
							E060101	Asystole	A0104	Migration or Expulsion of Device		
							E2007	Fall	A0104	Migration or Expulsion of Device		
							E2403	No Clinical Signs, Symptoms or Conditions	A0104	Migration or Expulsion of Device		
							F08	Hospitalization or Prolonged Hospitalization	A0104	Migration or Expulsion of Device		
							F1203	Life Threatening Illness or Injury	A0104	Migration or Expulsion of Device		
							F23	Unexpected Medical Intervention	A0104	Migration or Expulsion of Device		
					CT233	Surgical	E011903	Syncope/Fainting	A0104	Migration or Expulsion of Device		
							E060101	Asystole	A0104	Migration or Expulsion of Device		
							E2007	Fall	A0104	Migration or Expulsion of Device		
							E2403	No Clinical Signs, Symptoms or Conditions	A0104	Migration or Expulsion of Device		
							F08	Hospitalization or Prolonged Hospitalization	A0104	Migration or Expulsion of Device		
							F1203	Life Threatening Illness or Injury	A0104	Migration or Expulsion of Device		
							F23	Unexpected Medical Intervention	A0104	Migration or Expulsion of Device		
					CT2406	Implantable	E011903	Syncope/Fainting	A0104	Migration or Expulsion of Device		
							E060101	Asystole	A0104	Migration or Expulsion of Device		
							E2007	Fall	A0104	Migration or Expulsion of Device		
							E2403	No Clinical Signs, Symptoms or Conditions	A0104	Migration or Expulsion of Device		
							F08	Hospitalization or Prolonged Hospitalization	A0104	Migration or Expulsion of Device		
							F1203	Life Threatening Illness or Injury	A0104	Migration or Expulsion of Device		
							F23	Unexpected Medical Intervention	A0104	Migration or Expulsion of Device		
					CT2407	Active implantable	E011903	Syncope/Fainting	A0104	Migration or Expulsion of Device		
							E060101	Asystole	A0104	Migration or Expulsion of Device		
							E2007	Fall	A0104	Migration or Expulsion of Device		
							E2403	No Clinical Signs, Symptoms or Conditions	A0104	Migration or Expulsion of Device		
							F08	Hospitalization or Prolonged Hospitalization	A0104	Migration or Expulsion of Device		
							F1203	Life Threatening Illness or Injury	A0104	Migration or Expulsion of Device		
							F23	Unexpected Medical Intervention	A0104	Migration or Expulsion of Device		
					CT257	Implantable cardiac pacemakers	E011903	Syncope/Fainting	A0104	Migration or Expulsion of Device		
							E060101	Asystole	A0104	Migration or Expulsion of Device		
							E2007	Fall	A0104	Migration or Expulsion of Device		
							E2403	No Clinical Signs, Symptoms or Conditions	A0104	Migration or Expulsion of Device		
							F08	Hospitalization or Prolonged Hospitalization	A0104	Migration or Expulsion of Device		
							F1203	Life Threatening Illness or Injury	A0104	Migration or Expulsion of Device		
							F23	Unexpected Medical Intervention	A0104	Migration or Expulsion of Device		
					CT2727	Active implant devices/ancillaries	E011903	Syncope/Fainting	A0104	Migration or Expulsion of Device		
							E060101	Asystole	A0104	Migration or Expulsion of Device		
							E2007	Fall	A0104	Migration or Expulsion of Device		
							E2403	No Clinical Signs, Symptoms or Conditions	A0104	Migration or Expulsion of Device		
							F08	Hospitalization or Prolonged Hospitalization	A0104	Migration or Expulsion of Device		
							F1203	Life Threatening Illness or Injury	A0104	Migration or Expulsion of Device		
							F23	Unexpected Medical Intervention	A0104	Migration or Expulsion of Device		
					CT2750	Implantable cardiac electrical stimulators and associated devices	E011903	Syncope/Fainting	A0104	Migration or Expulsion of Device		
							E060101	Asystole	A0104	Migration or Expulsion of Device		
							E2007	Fall	A0104	Migration or Expulsion of Device		
							E2403	No Clinical Signs, Symptoms or Conditions	A0104	Migration or Expulsion of Device		
							F08	Hospitalization or Prolonged Hospitalization	A0104	Migration or Expulsion of Device		
							F1203	Life Threatening Illness or Injury	A0104	Migration or Expulsion of Device		
							F23	Unexpected Medical Intervention	A0104	Migration or Expulsion of Device		
					CT2866	Implantable cardiac electrical stimulators	E011903	Syncope/Fainting	A0104	Migration or Expulsion of Device		
							E060101	Asystole	A0104	Migration or Expulsion of Device		
							E2007	Fall	A0104	Migration or Expulsion of Device		
							E2403	No Clinical Signs, Symptoms or Conditions	A0104	Migration or Expulsion of Device		
							F08	Hospitalization or Prolonged Hospitalization	A0104	Migration or Expulsion of Device		

							F1203	Life Threatening Illness or Injury	A0104	Migration or Expulsion of Device		
							F23	Unexpected Medical Intervention	A0104	Migration or Expulsion of Device		
					CT2985	Active	E011903	Syncope/Fainting	A0104	Migration or Expulsion of Device		
							E060101	Asystole	A0104	Migration or Expulsion of Device		
							E2007	Fall	A0104	Migration or Expulsion of Device		
							E2403	No Clinical Signs, Symptoms or Conditions	A0104	Migration or Expulsion of Device		
							F08	Hospitalization or Prolonged Hospitalization	A0104	Migration or Expulsion of Device		
							F1203	Life Threatening Illness or Injury	A0104	Migration or Expulsion of Device		
							F23	Unexpected Medical Intervention	A0104	Migration or Expulsion of Device		
					CT3001	Implantable metal	E011903	Syncope/Fainting	A0104	Migration or Expulsion of Device		
							E060101	Asystole	A0104	Migration or Expulsion of Device		
							E2007	Fall	A0104	Migration or Expulsion of Device		
							E2403	No Clinical Signs, Symptoms or Conditions	A0104	Migration or Expulsion of Device		
							F08	Hospitalization or Prolonged Hospitalization	A0104	Migration or Expulsion of Device		
							F1203	Life Threatening Illness or Injury	A0104	Migration or Expulsion of Device		
							F23	Unexpected Medical Intervention	A0104	Migration or Expulsion of Device		
					CT3002	Implantable synthetic polymer	E011903	Syncope/Fainting	A0104	Migration or Expulsion of Device		
							E060101	Asystole	A0104	Migration or Expulsion of Device		
							E2007	Fall	A0104	Migration or Expulsion of Device		
							E2403	No Clinical Signs, Symptoms or Conditions	A0104	Migration or Expulsion of Device		
							F08	Hospitalization or Prolonged Hospitalization	A0104	Migration or Expulsion of Device		
							F1203	Life Threatening Illness or Injury	A0104	Migration or Expulsion of Device		
							F23	Unexpected Medical Intervention	A0104	Migration or Expulsion of Device		
					CT3006	Implantable inorganic material	E011903	Syncope/Fainting	A0104	Migration or Expulsion of Device		
							E060101	Asystole	A0104	Migration or Expulsion of Device		
							E2007	Fall	A0104	Migration or Expulsion of Device		
							E2403	No Clinical Signs, Symptoms or Conditions	A0104	Migration or Expulsion of Device		
							F08	Hospitalization or Prolonged Hospitalization	A0104	Migration or Expulsion of Device		
							F1203	Life Threatening Illness or Injury	A0104	Migration or Expulsion of Device		
							F23	Unexpected Medical Intervention	A0104	Migration or Expulsion of Device		
					CT314	Electrically powered	E011903	Syncope/Fainting	A0104	Migration or Expulsion of Device		
							E060101	Asystole	A0104	Migration or Expulsion of Device		
							E2007	Fall	A0104	Migration or Expulsion of Device		
							E2403	No Clinical Signs, Symptoms or Conditions	A0104	Migration or Expulsion of Device		
							F08	Hospitalization or Prolonged Hospitalization	A0104	Migration or Expulsion of Device		
							F1203	Life Threatening Illness or Injury	A0104	Migration or Expulsion of Device		
							F23	Unexpected Medical Intervention	A0104	Migration or Expulsion of Device		
					CT321	Long-term surgical invasive	E011903	Syncope/Fainting	A0104	Migration or Expulsion of Device		
							E060101	Asystole	A0104	Migration or Expulsion of Device		
							E2007	Fall	A0104	Migration or Expulsion of Device		
							E2403	No Clinical Signs, Symptoms or Conditions	A0104	Migration or Expulsion of Device		
							F08	Hospitalization or Prolonged Hospitalization	A0104	Migration or Expulsion of Device		
							F1203	Life Threatening Illness or Injury	A0104	Migration or Expulsion of Device		
							F23	Unexpected Medical Intervention	A0104	Migration or Expulsion of Device		
					CT449	Generators and associated devices	E011903	Syncope/Fainting	A0104	Migration or Expulsion of Device		
							E060101	Asystole	A0104	Migration or Expulsion of Device		
							E2007	Fall	A0104	Migration or Expulsion of Device		
							E2403	No Clinical Signs, Symptoms or Conditions	A0104	Migration or Expulsion of Device		
							F08	Hospitalization or Prolonged Hospitalization	A0104	Migration or Expulsion of Device		
							F1203	Life Threatening Illness or Injury	A0104	Migration or Expulsion of Device		
							F23	Unexpected Medical Intervention	A0104	Migration or Expulsion of Device		
					CT981	Single use	E011903	Syncope/Fainting	A0104	Migration or Expulsion of Device		
							E060101	Asystole	A0104	Migration or Expulsion of Device		
							E2007	Fall	A0104	Migration or Expulsion of Device		
							E2403	No Clinical Signs, Symptoms or Conditions	A0104	Migration or Expulsion of Device		
							F08	Hospitalization or Prolonged Hospitalization	A0104	Migration or Expulsion of Device		
							F1203	Life Threatening Illness or Injury	A0104	Migration or Expulsion of Device		
							F23	Unexpected Medical Intervention	A0104	Migration or Expulsion of Device		
					CT983	Surgical invasive	E011903	Syncope/Fainting	A0104	Migration or Expulsion of Device		
							E060101	Asystole	A0104	Migration or Expulsion of Device		
							E2007	Fall	A0104	Migration or Expulsion of Device		
							E2403	No Clinical Signs, Symptoms or Conditions	A0104	Migration or Expulsion of Device		
							F08	Hospitalization or Prolonged Hospitalization	A0104	Migration or Expulsion of Device		
							F1203	Life Threatening Illness or Injury	A0104	Migration or Expulsion of Device		
							F23	Unexpected Medical Intervention	A0104	Migration or Expulsion of Device		
7754	2022	All other reportable incidents	35223	Endocardial/interventricular septal pacing lead	CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A0401	Break		
									A070909	Oversensing		
									A0722	Impedance Problem		
							F26	No Health Consequences or Impact	A0401	Break		
									A070909	Oversensing		
									A0722	Impedance Problem		
7755	2022	All other reportable incidents	47265	Dual-chamber implantable pacemaker, rate-responsive	CT141	Electrodes/Leads and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F12	Serious Injury/ Illness/ Impairment	A070504	Premature Discharge of Battery		
					CT145	Cardiovascular devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F12	Serious Injury/ Illness/ Impairment	A070504	Premature Discharge of Battery		
					CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F12	Serious Injury/ Illness/ Impairment	A070504	Premature Discharge of Battery		
					CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F12	Serious Injury/ Illness/ Impairment	A070504	Premature Discharge of Battery		
					CT233	Surgical	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F12	Serious Injury/ Illness/ Impairment	A070504	Premature Discharge of Battery		
					CT2406	Implantable	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F12	Serious Injury/ Illness/ Impairment	A070504	Premature Discharge of Battery		
					CT2407	Active implantable	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F12	Serious Injury/ Illness/ Impairment	A070504	Premature Discharge of Battery		
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F12	Serious Injury/ Illness/ Impairment	A070504	Premature Discharge of Battery		
					CT2727	Active implant devices/ancillaries	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		

							F12	Serious Injury/ Illness/ Impairment	A070504	Premature Discharge of Battery		
					CT2750	Implantable cardiac electrical stimulators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F12	Serious Injury/ Illness/ Impairment	A070504	Premature Discharge of Battery		
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F12	Serious Injury/ Illness/ Impairment	A070504	Premature Discharge of Battery		
					CT2985	Active	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F12	Serious Injury/ Illness/ Impairment	A070504	Premature Discharge of Battery		
					CT3001	Implantable metal	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F12	Serious Injury/ Illness/ Impairment	A070504	Premature Discharge of Battery		
					CT3002	Implantable synthetic polymer	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F12	Serious Injury/ Illness/ Impairment	A070504	Premature Discharge of Battery		
					CT3006	Implantable inorganic material	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F12	Serious Injury/ Illness/ Impairment	A070504	Premature Discharge of Battery		
					CT314	Electrically powered	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F12	Serious Injury/ Illness/ Impairment	A070504	Premature Discharge of Battery		
					CT321	Long-term surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F12	Serious Injury/ Illness/ Impairment	A070504	Premature Discharge of Battery		
					CT449	Generators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F12	Serious Injury/ Illness/ Impairment	A070504	Premature Discharge of Battery		
					CT981	Single use	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F12	Serious Injury/ Illness/ Impairment	A070504	Premature Discharge of Battery		
					CT983	Surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F12	Serious Injury/ Illness/ Impairment	A070504	Premature Discharge of Battery		
7756	2022	All other reportable incidents	35223	Endocardial/interventricular septal pacing lead	CT1684	Cardiac pacemakers and associated devices	E0717	Dyspnea	A051201	Device Dislodged or Dislocated	27/01/2022	27/01/2022
									A070102	High Capture Threshold	27/01/2022	27/01/2022
							E2403	No Clinical Signs, Symptoms or Conditions	A051201	Device Dislodged or Dislocated	27/01/2022	27/01/2022
									A070102	High Capture Threshold	27/01/2022	27/01/2022
							F08	Hospitalization or Prolonged Hospitalization	A051201	Device Dislodged or Dislocated	27/01/2022	27/01/2022
									A070102	High Capture Threshold	27/01/2022	27/01/2022
							F11	Non-Serious Injury/ Illness/ Impairment	A051201	Device Dislodged or Dislocated	27/01/2022	27/01/2022
									A070102	High Capture Threshold	27/01/2022	27/01/2022
							F12	Serious Injury/ Illness/ Impairment	A051201	Device Dislodged or Dislocated	27/01/2022	27/01/2022
									A070102	High Capture Threshold	27/01/2022	27/01/2022
							F1905	Device Revision or Replacement	A051201	Device Dislodged or Dislocated	27/01/2022	27/01/2022
									A070102	High Capture Threshold	27/01/2022	27/01/2022
7757	2022	All other reportable incidents	47270	Cardiac resynchronization therapy implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A0713	Defibrillation/Stimulation Problem		
							F26	No Health Consequences or Impact	A0713	Defibrillation/Stimulation Problem		
7758	2022	All other reportable incidents	37265	Dual-chamber implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A072201	High Impedance		
							F26	No Health Consequences or Impact	A072201	High Impedance		
7759	2022	All other reportable incidents	35852	Single-chamber implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E2104	Electric Shock	A010402	Migration		
									A070909	Oversensing		
									A071302	Inappropriate/Inadequate Shock/Stimulation		
									A1301	Failure to Read Input Signal		
							E2117	Twiddlers Syndrome	A010402	Migration		
									A070909	Oversensing		
									A071302	Inappropriate/Inadequate Shock/Stimulation		
									A1301	Failure to Read Input Signal		
							E2403	No Clinical Signs, Symptoms or Conditions	A010402	Migration		
									A070909	Oversensing		
									A071302	Inappropriate/Inadequate Shock/Stimulation		
									A1301	Failure to Read Input Signal		
							F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A010402	Migration		
									A070909	Oversensing		
									A071302	Inappropriate/Inadequate Shock/Stimulation		
									A1301	Failure to Read Input Signal		
7760	2022	All other reportable incidents	47263	Cardiac resynchronization therapy implantable pacemaker	CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem	29/03/2022	29/03/2022
									A070504	Premature Discharge of Battery	29/03/2022	29/03/2022
							F08	Hospitalization or Prolonged Hospitalization	A0705	Battery Problem	29/03/2022	29/03/2022
									A070504	Premature Discharge of Battery	29/03/2022	29/03/2022
							F1905	Device Revision or Replacement	A0705	Battery Problem	29/03/2022	29/03/2022
									A070504	Premature Discharge of Battery	29/03/2022	29/03/2022
					CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem	29/03/2022	29/03/2022
									A070504	Premature Discharge of Battery	29/03/2022	29/03/2022
							F08	Hospitalization or Prolonged Hospitalization	A0705	Battery Problem	29/03/2022	29/03/2022
									A070504	Premature Discharge of Battery	29/03/2022	29/03/2022
							F1905	Device Revision or Replacement	A0705	Battery Problem	29/03/2022	29/03/2022
									A070504	Premature Discharge of Battery	29/03/2022	29/03/2022
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem	29/03/2022	29/03/2022
									A070504	Premature Discharge of Battery	29/03/2022	29/03/2022
							F08	Hospitalization or Prolonged Hospitalization	A0705	Battery Problem	29/03/2022	29/03/2022
									A070504	Premature Discharge of Battery	29/03/2022	29/03/2022
							F1905	Device Revision or Replacement	A0705	Battery Problem	29/03/2022	29/03/2022
									A070504	Premature Discharge of Battery	29/03/2022	29/03/2022
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem	29/03/2022	29/03/2022
									A070504	Premature Discharge of Battery	29/03/2022	29/03/2022
							F08	Hospitalization or Prolonged Hospitalization	A0705	Battery Problem	29/03/2022	29/03/2022
									A070504	Premature Discharge of Battery	29/03/2022	29/03/2022
							F1905	Device Revision or Replacement	A0705	Battery Problem	29/03/2022	29/03/2022
									A070504	Premature Discharge of Battery	29/03/2022	29/03/2022
7761	2022	All other reportable incidents	36241	Pacemaker/defibrillator lead	CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A0203	Defective Device		
									A070909	Oversensing		
							F23	Unexpected Medical Intervention	A0203	Defective Device		
									A070909	Oversensing		
7762	2022	All other reportable incidents	17882	Physiologic-monitoring defibrillation system	CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070908	Failure to Sense		
							F26	No Health Consequences or Impact	A070908	Failure to Sense		
7763	2022	All other reportable incidents	47265	Dual-chamber implantable pacemaker, rate-responsive	CT141	Electrodes/Leads and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	04/01/2022	04/01/2022
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	04/01/2022	04/01/2022
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	04/01/2022	04/01/2022

					CT145	Cardiovascular devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	04/01/2022	04/01/2022
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	04/01/2022	04/01/2022
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	04/01/2022	04/01/2022
					CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	04/01/2022	04/01/2022
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	04/01/2022	04/01/2022
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	04/01/2022	04/01/2022
					CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	04/01/2022	04/01/2022
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	04/01/2022	04/01/2022
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	04/01/2022	04/01/2022
					CT233	Surgical	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	04/01/2022	04/01/2022
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	04/01/2022	04/01/2022
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	04/01/2022	04/01/2022
					CT2406	Implantable	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	04/01/2022	04/01/2022
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	04/01/2022	04/01/2022
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	04/01/2022	04/01/2022
					CT2407	Active implantable	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	04/01/2022	04/01/2022
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	04/01/2022	04/01/2022
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	04/01/2022	04/01/2022
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	04/01/2022	04/01/2022
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	04/01/2022	04/01/2022
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	04/01/2022	04/01/2022
					CT2727	Active implant devices/ancillaries	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	04/01/2022	04/01/2022
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	04/01/2022	04/01/2022
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	04/01/2022	04/01/2022
					CT2750	Implantable cardiac electrical stimulators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	04/01/2022	04/01/2022
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	04/01/2022	04/01/2022
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	04/01/2022	04/01/2022
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	04/01/2022	04/01/2022
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	04/01/2022	04/01/2022
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	04/01/2022	04/01/2022
					CT2985	Active	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	04/01/2022	04/01/2022
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	04/01/2022	04/01/2022
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	04/01/2022	04/01/2022
					CT3001	Implantable metal	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	04/01/2022	04/01/2022
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	04/01/2022	04/01/2022
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	04/01/2022	04/01/2022
					CT3002	Implantable synthetic polymer	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	04/01/2022	04/01/2022
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	04/01/2022	04/01/2022
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	04/01/2022	04/01/2022
					CT3006	Implantable inorganic material	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	04/01/2022	04/01/2022
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	04/01/2022	04/01/2022
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	04/01/2022	04/01/2022
					CT314	Electrically powered	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	04/01/2022	04/01/2022
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	04/01/2022	04/01/2022
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	04/01/2022	04/01/2022
					CT321	Long-term surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	04/01/2022	04/01/2022
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	04/01/2022	04/01/2022
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	04/01/2022	04/01/2022
					CT449	Generators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	04/01/2022	04/01/2022
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	04/01/2022	04/01/2022
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	04/01/2022	04/01/2022
					CT981	Single use	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	04/01/2022	04/01/2022
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	04/01/2022	04/01/2022
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	04/01/2022	04/01/2022
					CT983	Surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	04/01/2022	04/01/2022
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	04/01/2022	04/01/2022
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	04/01/2022	04/01/2022
7764	2022	All other reportable incidents	35852	Single-chamber implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E2104	Electric Shock	A071302	Inappropriate/Inadequate Shock/Stimulation		
							E2403	No Clinical Signs, Symptoms or Conditions	A071302	Inappropriate/Inadequate Shock/Stimulation		
							F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A071302	Inappropriate/Inadequate Shock/Stimulation		
7765	2022	All other reportable incidents	35852	Single-chamber implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A072201	High Impedance		
							F26	No Health Consequences or Impact	A072201	High Impedance		
7766	2022	All other reportable incidents	47265	Dual-chamber implantable pacemaker, rate-responsive	CT141	Electrodes/Leads and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	15/08/2022	15/08/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	15/08/2022	15/08/2022
					CT145	Cardiovascular devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	15/08/2022	15/08/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	15/08/2022	15/08/2022
					CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	15/08/2022	15/08/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	15/08/2022	15/08/2022
					CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	15/08/2022	15/08/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	15/08/2022	15/08/2022
					CT233	Surgical	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	15/08/2022	15/08/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	15/08/2022	15/08/2022
					CT2406	Implantable	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	15/08/2022	15/08/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	15/08/2022	15/08/2022
					CT2407	Active implantable	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	15/08/2022	15/08/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	15/08/2022	15/08/2022
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	15/08/2022	15/08/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	15/08/2022	15/08/2022
					CT2727	Active implant devices/ancillaries	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	15/08/2022	15/08/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	15/08/2022	15/08/2022
					CT2750	Implantable cardiac electrical stimulators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	15/08/2022	15/08/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	15/08/2022	15/08/2022
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	15/08/2022	15/08/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	15/08/2022	15/08/2022
					CT2985	Active	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	15/08/2022	15/08/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	15/08/2022	15/08/2022
					CT3001	Implantable metal	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	15/08/2022	15/08/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	15/08/2022	15/08/2022

					CT3002	Implantable synthetic polymer	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	15/08/2022	15/08/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	15/08/2022	15/08/2022
					CT3006	Implantable inorganic material	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	15/08/2022	15/08/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	15/08/2022	15/08/2022
					CT314	Electrically powered	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	15/08/2022	15/08/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	15/08/2022	15/08/2022
					CT321	Long-term surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	15/08/2022	15/08/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	15/08/2022	15/08/2022
					CT449	Generators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	15/08/2022	15/08/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	15/08/2022	15/08/2022
					CT981	Single use	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	15/08/2022	15/08/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	15/08/2022	15/08/2022
					CT983	Surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	15/08/2022	15/08/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	15/08/2022	15/08/2022
7767	2022	All other reportable incidents	47265	Dual-chamber implantable pacemaker, rate-responsive	CT141	Electrodes/Leads and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070101	Failure to Capture		
									A0712	Pacing Problem		
							F23	Unexpected Medical Intervention	A070101	Failure to Capture		
									A0712	Pacing Problem		
					CT145	Cardiovascular devices	E2403	No Clinical Signs, Symptoms or Conditions	A070101	Failure to Capture		
									A0712	Pacing Problem		
							F23	Unexpected Medical Intervention	A070101	Failure to Capture		
									A0712	Pacing Problem		
					CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070101	Failure to Capture		
									A0712	Pacing Problem		
							F23	Unexpected Medical Intervention	A070101	Failure to Capture		
									A0712	Pacing Problem		
					CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A070101	Failure to Capture		
									A0712	Pacing Problem		
							F23	Unexpected Medical Intervention	A070101	Failure to Capture		
									A0712	Pacing Problem		
					CT233	Surgical	E2403	No Clinical Signs, Symptoms or Conditions	A070101	Failure to Capture		
									A0712	Pacing Problem		
							F23	Unexpected Medical Intervention	A070101	Failure to Capture		
									A0712	Pacing Problem		
					CT2406	Implantable	E2403	No Clinical Signs, Symptoms or Conditions	A070101	Failure to Capture		
									A0712	Pacing Problem		
							F23	Unexpected Medical Intervention	A070101	Failure to Capture		
									A0712	Pacing Problem		
					CT2407	Active implantable	E2403	No Clinical Signs, Symptoms or Conditions	A070101	Failure to Capture		
									A0712	Pacing Problem		
							F23	Unexpected Medical Intervention	A070101	Failure to Capture		
									A0712	Pacing Problem		
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A070101	Failure to Capture		
									A0712	Pacing Problem		
							F23	Unexpected Medical Intervention	A070101	Failure to Capture		
									A0712	Pacing Problem		
					CT2727	Active implant devices/ancillaries	E2403	No Clinical Signs, Symptoms or Conditions	A070101	Failure to Capture		
									A0712	Pacing Problem		
							F23	Unexpected Medical Intervention	A070101	Failure to Capture		
									A0712	Pacing Problem		
					CT2750	Implantable cardiac electrical stimulators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070101	Failure to Capture		
									A0712	Pacing Problem		
							F23	Unexpected Medical Intervention	A070101	Failure to Capture		
									A0712	Pacing Problem		
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A070101	Failure to Capture		
									A0712	Pacing Problem		
							F23	Unexpected Medical Intervention	A070101	Failure to Capture		
									A0712	Pacing Problem		
					CT2985	Active	E2403	No Clinical Signs, Symptoms or Conditions	A070101	Failure to Capture		
									A0712	Pacing Problem		
							F23	Unexpected Medical Intervention	A070101	Failure to Capture		
									A0712	Pacing Problem		
					CT3001	Implantable metal	E2403	No Clinical Signs, Symptoms or Conditions	A070101	Failure to Capture		
									A0712	Pacing Problem		
							F23	Unexpected Medical Intervention	A070101	Failure to Capture		
									A0712	Pacing Problem		
					CT3002	Implantable synthetic polymer	E2403	No Clinical Signs, Symptoms or Conditions	A070101	Failure to Capture		
									A0712	Pacing Problem		
							F23	Unexpected Medical Intervention	A070101	Failure to Capture		
									A0712	Pacing Problem		
					CT3006	Implantable inorganic material	E2403	No Clinical Signs, Symptoms or Conditions	A070101	Failure to Capture		
									A0712	Pacing Problem		
							F23	Unexpected Medical Intervention	A070101	Failure to Capture		
									A0712	Pacing Problem		
					CT314	Electrically powered	E2403	No Clinical Signs, Symptoms or Conditions	A070101	Failure to Capture		
									A0712	Pacing Problem		
							F23	Unexpected Medical Intervention	A070101	Failure to Capture		
									A0712	Pacing Problem		
					CT321	Long-term surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A070101	Failure to Capture		
									A0712	Pacing Problem		
							F23	Unexpected Medical Intervention	A070101	Failure to Capture		
									A0712	Pacing Problem		
					CT449	Generators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070101	Failure to Capture		
									A0712	Pacing Problem		
							F23	Unexpected Medical Intervention	A070101	Failure to Capture		
									A0712	Pacing Problem		
					CT981	Single use	E2403	No Clinical Signs, Symptoms or Conditions	A070101	Failure to Capture		
									A0712	Pacing Problem		

							F23	Unexpected Medical Intervention	A070101	Failure to Capture		
									A0712	Pacing Problem		
					CT983	Surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A070101	Failure to Capture		
									A0712	Pacing Problem		
							F23	Unexpected Medical Intervention	A070101	Failure to Capture		
									A0712	Pacing Problem		
7768	2022	All other reportable incidents	47265	Dual-chamber implantable pacemaker, rate-responsive	CT141	Electrodes/Leads and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	02/09/2022	02/09/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	02/09/2022	02/09/2022
					CT145	Cardiovascular devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	02/09/2022	02/09/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	02/09/2022	02/09/2022
					CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	02/09/2022	02/09/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	02/09/2022	02/09/2022
					CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	02/09/2022	02/09/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	02/09/2022	02/09/2022
					CT233	Surgical	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	02/09/2022	02/09/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	02/09/2022	02/09/2022
					CT2406	Implantable	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	02/09/2022	02/09/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	02/09/2022	02/09/2022
					CT2407	Active implantable	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	02/09/2022	02/09/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	02/09/2022	02/09/2022
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	02/09/2022	02/09/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	02/09/2022	02/09/2022
					CT2727	Active implant devices/ancillaries	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	02/09/2022	02/09/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	02/09/2022	02/09/2022
					CT2750	Implantable cardiac electrical stimulators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	02/09/2022	02/09/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	02/09/2022	02/09/2022
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	02/09/2022	02/09/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	02/09/2022	02/09/2022
					CT2985	Active	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	02/09/2022	02/09/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	02/09/2022	02/09/2022
					CT3001	Implantable metal	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	02/09/2022	02/09/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	02/09/2022	02/09/2022
					CT3002	Implantable synthetic polymer	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	02/09/2022	02/09/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	02/09/2022	02/09/2022
					CT3006	Implantable inorganic material	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	02/09/2022	02/09/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	02/09/2022	02/09/2022
					CT314	Electrically powered	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	02/09/2022	02/09/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	02/09/2022	02/09/2022
					CT321	Long-term surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	02/09/2022	02/09/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	02/09/2022	02/09/2022
					CT449	Generators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	02/09/2022	02/09/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	02/09/2022	02/09/2022
					CT981	Single use	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	02/09/2022	02/09/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	02/09/2022	02/09/2022
					CT983	Surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	02/09/2022	02/09/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	02/09/2022	02/09/2022
7769	2022		37265	Dual-chamber implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	16/09/2022	16/09/2022
							F1901	Additional Surgery	A070504	Premature Discharge of Battery	16/09/2022	16/09/2022
7770	2022	All other reportable incidents	35852	Single-chamber implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A072201	High Impedance		
							F2203	Imaging Required	A072201	High Impedance		
							F26	No Health Consequences or Impact	A072201	High Impedance		
7771	2022	All other reportable incidents	35852	Single-chamber implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E0601	Arrhythmia	A070910	Undersensing		
									A0909	Unexpected Therapeutic Results		
							E060110	Ventricular Fibrillation	A070910	Undersensing		
									A0909	Unexpected Therapeutic Results		
							E2403	No Clinical Signs, Symptoms or Conditions	A070910	Undersensing		
									A0909	Unexpected Therapeutic Results		
							F26	No Health Consequences or Impact	A070910	Undersensing		
									A0909	Unexpected Therapeutic Results		
7772	2022	All other reportable incidents	35223	Endocardial/interventricular septal pacing lead	CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070909	Oversensing		
							F23	Unexpected Medical Intervention	A070909	Oversensing		
7773	2022	All other reportable incidents	47265	Dual-chamber implantable pacemaker, rate-responsive	CT141	Electrodes/Leads and associated devices	E0601	Arrhythmia	A070101	Failure to Capture		
							E2403	No Clinical Signs, Symptoms or Conditions	A070101	Failure to Capture		
							F11	Non-Serious Injury/ Illness/ Impairment	A070101	Failure to Capture		
					CT145	Cardiovascular devices	E0601	Arrhythmia	A070101	Failure to Capture		
							E2403	No Clinical Signs, Symptoms or Conditions	A070101	Failure to Capture		
							F11	Non-Serious Injury/ Illness/ Impairment	A070101	Failure to Capture		
					CT1684	Cardiac pacemakers and associated devices	E0601	Arrhythmia	A070101	Failure to Capture		
							E2403	No Clinical Signs, Symptoms or Conditions	A070101	Failure to Capture		
							F11	Non-Serious Injury/ Illness/ Impairment	A070101	Failure to Capture		
					CT1685	Cardiac pacemakers	E0601	Arrhythmia	A070101	Failure to Capture		
							E2403	No Clinical Signs, Symptoms or Conditions	A070101	Failure to Capture		
							F11	Non-Serious Injury/ Illness/ Impairment	A070101	Failure to Capture		
					CT233	Surgical	E0601	Arrhythmia	A070101	Failure to Capture		
							E2403	No Clinical Signs, Symptoms or Conditions	A070101	Failure to Capture		
							F11	Non-Serious Injury/ Illness/ Impairment	A070101	Failure to Capture		
					CT2406	Implantable	E0601	Arrhythmia	A070101	Failure to Capture		
							E2403	No Clinical Signs, Symptoms or Conditions	A070101	Failure to Capture		
							F11	Non-Serious Injury/ Illness/ Impairment	A070101	Failure to Capture		
					CT2407	Active implantable	E0601	Arrhythmia	A070101	Failure to Capture		
							E2403	No Clinical Signs, Symptoms or Conditions	A070101	Failure to Capture		
							F11	Non-Serious Injury/ Illness/ Impairment	A070101	Failure to Capture		
					CT257	Implantable cardiac pacemakers	E0601	Arrhythmia	A070101	Failure to Capture		
							E2403	No Clinical Signs, Symptoms or Conditions	A070101	Failure to Capture		
							F11	Non-Serious Injury/ Illness/ Impairment	A070101	Failure to Capture		
					CT2727	Active implant devices/ancillaries	E0601	Arrhythmia	A070101	Failure to Capture		
							E2403	No Clinical Signs, Symptoms or Conditions	A070101	Failure to Capture		
							F11	Non-Serious Injury/ Illness/ Impairment	A070101	Failure to Capture		

					CT2750	Implantable cardiac electrical stimulators and associated devices	E0601	Arrhythmia	A070101	Failure to Capture		
							E2403	No Clinical Signs, Symptoms or Conditions	A070101	Failure to Capture		
							F11	Non-Serious Injury/ Illness/ Impairment	A070101	Failure to Capture		
					CT2866	Implantable cardiac electrical stimulators	E0601	Arrhythmia	A070101	Failure to Capture		
							E2403	No Clinical Signs, Symptoms or Conditions	A070101	Failure to Capture		
							F11	Non-Serious Injury/ Illness/ Impairment	A070101	Failure to Capture		
					CT2985	Active	E0601	Arrhythmia	A070101	Failure to Capture		
							E2403	No Clinical Signs, Symptoms or Conditions	A070101	Failure to Capture		
							F11	Non-Serious Injury/ Illness/ Impairment	A070101	Failure to Capture		
					CT3001	Implantable metal	E0601	Arrhythmia	A070101	Failure to Capture		
							E2403	No Clinical Signs, Symptoms or Conditions	A070101	Failure to Capture		
							F11	Non-Serious Injury/ Illness/ Impairment	A070101	Failure to Capture		
					CT3002	Implantable synthetic polymer	E0601	Arrhythmia	A070101	Failure to Capture		
							E2403	No Clinical Signs, Symptoms or Conditions	A070101	Failure to Capture		
							F11	Non-Serious Injury/ Illness/ Impairment	A070101	Failure to Capture		
					CT3006	Implantable inorganic material	E0601	Arrhythmia	A070101	Failure to Capture		
							E2403	No Clinical Signs, Symptoms or Conditions	A070101	Failure to Capture		
							F11	Non-Serious Injury/ Illness/ Impairment	A070101	Failure to Capture		
					CT314	Electrically powered	E0601	Arrhythmia	A070101	Failure to Capture		
							E2403	No Clinical Signs, Symptoms or Conditions	A070101	Failure to Capture		
							F11	Non-Serious Injury/ Illness/ Impairment	A070101	Failure to Capture		
					CT321	Long-term surgical invasive	E0601	Arrhythmia	A070101	Failure to Capture		
							E2403	No Clinical Signs, Symptoms or Conditions	A070101	Failure to Capture		
							F11	Non-Serious Injury/ Illness/ Impairment	A070101	Failure to Capture		
					CT449	Generators and associated devices	E0601	Arrhythmia	A070101	Failure to Capture		
							E2403	No Clinical Signs, Symptoms or Conditions	A070101	Failure to Capture		
							F11	Non-Serious Injury/ Illness/ Impairment	A070101	Failure to Capture		
					CT981	Single use	E0601	Arrhythmia	A070101	Failure to Capture		
							E2403	No Clinical Signs, Symptoms or Conditions	A070101	Failure to Capture		
							F11	Non-Serious Injury/ Illness/ Impairment	A070101	Failure to Capture		
					CT983	Surgical invasive	E0601	Arrhythmia	A070101	Failure to Capture		
							E2403	No Clinical Signs, Symptoms or Conditions	A070101	Failure to Capture		
							F11	Non-Serious Injury/ Illness/ Impairment	A070101	Failure to Capture		
7774	2022	All other reportable incidents	47265	Dual-chamber implantable pacemaker, rate-responsive	CT141	Electrodes/Leads and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery		
							F12	Serious Injury/ Illness/ Impairment	A070504	Premature Discharge of Battery		
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery		
							F23	Unexpected Medical Intervention	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT145	Cardiovascular devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery		
							F12	Serious Injury/ Illness/ Impairment	A070504	Premature Discharge of Battery		
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery		
							F23	Unexpected Medical Intervention	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery		
							F12	Serious Injury/ Illness/ Impairment	A070504	Premature Discharge of Battery		
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery		
							F23	Unexpected Medical Intervention	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery		
							F12	Serious Injury/ Illness/ Impairment	A070504	Premature Discharge of Battery		
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery		
							F23	Unexpected Medical Intervention	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT233	Surgical	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery		
							F12	Serious Injury/ Illness/ Impairment	A070504	Premature Discharge of Battery		
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery		
							F23	Unexpected Medical Intervention	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT2406	Implantable	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery		
							F12	Serious Injury/ Illness/ Impairment	A070504	Premature Discharge of Battery		
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery		
							F23	Unexpected Medical Intervention	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT2407	Active implantable	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery		
							F12	Serious Injury/ Illness/ Impairment	A070504	Premature Discharge of Battery		
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery		
							F23	Unexpected Medical Intervention	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery		
							F12	Serious Injury/ Illness/ Impairment	A070504	Premature Discharge of Battery		
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery		
							F23	Unexpected Medical Intervention	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT2727	Active implant devices/ancillaries	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery		
							F12	Serious Injury/ Illness/ Impairment	A070504	Premature Discharge of Battery		
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery		
							F23	Unexpected Medical Intervention	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT2750	Implantable cardiac electrical stimulators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		

							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery		
							F12	Serious Injury/ Illness/ Impairment	A070504	Premature Discharge of Battery		
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery		
							F23	Unexpected Medical Intervention	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery		
							F12	Serious Injury/ Illness/ Impairment	A070504	Premature Discharge of Battery		
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery		
							F23	Unexpected Medical Intervention	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT2985	Active	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery		
							F12	Serious Injury/ Illness/ Impairment	A070504	Premature Discharge of Battery		
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery		
							F23	Unexpected Medical Intervention	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT3001	Implantable metal	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery		
							F12	Serious Injury/ Illness/ Impairment	A070504	Premature Discharge of Battery		
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery		
							F23	Unexpected Medical Intervention	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT3002	Implantable synthetic polymer	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery		
							F12	Serious Injury/ Illness/ Impairment	A070504	Premature Discharge of Battery		
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery		
							F23	Unexpected Medical Intervention	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT3006	Implantable inorganic material	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery		
							F12	Serious Injury/ Illness/ Impairment	A070504	Premature Discharge of Battery		
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery		
							F23	Unexpected Medical Intervention	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT314	Electrically powered	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery		
							F12	Serious Injury/ Illness/ Impairment	A070504	Premature Discharge of Battery		
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery		
							F23	Unexpected Medical Intervention	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT321	Long-term surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery		
							F12	Serious Injury/ Illness/ Impairment	A070504	Premature Discharge of Battery		
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery		
							F23	Unexpected Medical Intervention	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT449	Generators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery		
							F12	Serious Injury/ Illness/ Impairment	A070504	Premature Discharge of Battery		
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery		
							F23	Unexpected Medical Intervention	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT981	Single use	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery		
							F12	Serious Injury/ Illness/ Impairment	A070504	Premature Discharge of Battery		
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery		
							F23	Unexpected Medical Intervention	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT983	Surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery		
							F12	Serious Injury/ Illness/ Impairment	A070504	Premature Discharge of Battery		
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery		
							F23	Unexpected Medical Intervention	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
7775	2022	All other reportable incidents	17882	Physiologic-monitoring defibrillation system	CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A07	Electrical /Electronic Property Problem		
							F27	Problem Identified During Non-Clinical Procedure	A07	Electrical /Electronic Property Problem		
7776	2022	All other reportable incidents	47265	Dual-chamber implantable pacemaker, rate-responsive	CT141	Electrodes/Leads and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT145	Cardiovascular devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT233	Surgical	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT2406	Implantable	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT2407	Active implantable	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT2727	Active implant devices/ancillaries	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT2750	Implantable cardiac electrical stimulators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		

							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT2985	Active	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT3001	Implantable metal	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT3002	Implantable synthetic polymer	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT3006	Implantable inorganic material	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT314	Electrically powered	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT321	Long-term surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT449	Generators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT981	Single use	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT983	Surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
7777	2022	All other reportable incidents	17882	Physiologic-monitoring defibrillation system	CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A090811	Unable to Obtain Readings		
							F27	Problem Identified During Non-Clinical Procedure	A090811	Unable to Obtain Readings		
7778	2022	All other reportable incidents	47270	Cardiac resynchronization therapy implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A072201	High Impedance		
							F26	No Health Consequences or Impact	A072201	High Impedance		
7779	2022	All other reportable incidents	47265	Dual-chamber implantable pacemaker, rate-responsive	CT141	Electrodes/Leads and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
					CT145	Cardiovascular devices	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
					CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
					CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
					CT233	Surgical	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
					CT2406	Implantable	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
					CT2407	Active implantable	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
					CT2727	Active implant devices/ancillaries	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
					CT2750	Implantable cardiac electrical stimulators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
					CT2985	Active	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
					CT3001	Implantable metal	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
					CT3002	Implantable synthetic polymer	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
					CT3006	Implantable inorganic material	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
					CT314	Electrically powered	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
					CT321	Long-term surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem		

								A070504	Premature Discharge of Battery		
						F26	No Health Consequences or Impact	A0705	Battery Problem		
								A070504	Premature Discharge of Battery		
					CT449	Generators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
					CT981	Single use	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
					CT983	Surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
7780	2022	All other reportable incidents	36241	Pacemaker/defibrillator lead	CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A072201	High Impedance		
							F26	No Health Consequences or Impact	A072201	High Impedance		
7781	2022		47265	Dual-chamber implantable pacemaker, rate-responsive	CT141	Electrodes/Leads and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	14/09/2022	14/09/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	14/09/2022	14/09/2022
					CT145	Cardiovascular devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	14/09/2022	14/09/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	14/09/2022	14/09/2022
					CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	14/09/2022	14/09/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	14/09/2022	14/09/2022
					CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	14/09/2022	14/09/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	14/09/2022	14/09/2022
					CT233	Surgical	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	14/09/2022	14/09/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	14/09/2022	14/09/2022
					CT2406	Implantable	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	14/09/2022	14/09/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	14/09/2022	14/09/2022
					CT2407	Active implantable	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	14/09/2022	14/09/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	14/09/2022	14/09/2022
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	14/09/2022	14/09/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	14/09/2022	14/09/2022
					CT2727	Active implant devices/ancillaries	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	14/09/2022	14/09/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	14/09/2022	14/09/2022
					CT2750	Implantable cardiac electrical stimulators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	14/09/2022	14/09/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	14/09/2022	14/09/2022
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	14/09/2022	14/09/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	14/09/2022	14/09/2022
					CT2985	Active	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	14/09/2022	14/09/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	14/09/2022	14/09/2022
					CT3001	Implantable metal	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	14/09/2022	14/09/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	14/09/2022	14/09/2022
					CT3002	Implantable synthetic polymer	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	14/09/2022	14/09/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	14/09/2022	14/09/2022
					CT3006	Implantable inorganic material	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	14/09/2022	14/09/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	14/09/2022	14/09/2022
					CT314	Electrically powered	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	14/09/2022	14/09/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	14/09/2022	14/09/2022
					CT321	Long-term surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	14/09/2022	14/09/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	14/09/2022	14/09/2022
					CT449	Generators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	14/09/2022	14/09/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	14/09/2022	14/09/2022
					CT981	Single use	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	14/09/2022	14/09/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	14/09/2022	14/09/2022
					CT983	Surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	14/09/2022	14/09/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	14/09/2022	14/09/2022
7782	2022	All other reportable incidents	47265	Dual-chamber implantable pacemaker, rate-responsive	CT141	Electrodes/Leads and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	14/09/2022	14/09/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	14/09/2022	14/09/2022
					CT145	Cardiovascular devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	14/09/2022	14/09/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	14/09/2022	14/09/2022
					CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	14/09/2022	14/09/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	14/09/2022	14/09/2022
					CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	14/09/2022	14/09/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	14/09/2022	14/09/2022
					CT233	Surgical	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	14/09/2022	14/09/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	14/09/2022	14/09/2022
					CT2406	Implantable	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	14/09/2022	14/09/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	14/09/2022	14/09/2022
					CT2407	Active implantable	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	14/09/2022	14/09/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	14/09/2022	14/09/2022
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	14/09/2022	14/09/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	14/09/2022	14/09/2022
					CT2727	Active implant devices/ancillaries	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	14/09/2022	14/09/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	14/09/2022	14/09/2022
					CT2750	Implantable cardiac electrical stimulators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	14/09/2022	14/09/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	14/09/2022	14/09/2022
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	14/09/2022	14/09/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	14/09/2022	14/09/2022
					CT2985	Active	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	14/09/2022	14/09/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	14/09/2022	14/09/2022
					CT3001	Implantable metal	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	14/09/2022	14/09/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	14/09/2022	14/09/2022
					CT3002	Implantable synthetic polymer	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	14/09/2022	14/09/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	14/09/2022	14/09/2022
					CT3006	Implantable inorganic material	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	14/09/2022	14/09/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	14/09/2022	14/09/2022
					CT314	Electrically powered	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	14/09/2022	14/09/2022

							F1901	Additional Surgery	A24	No Apparent Device Problem	14/09/2022	14/09/2022
					CT321	Long-term surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	14/09/2022	14/09/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	14/09/2022	14/09/2022
					CT449	Generators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	14/09/2022	14/09/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	14/09/2022	14/09/2022
					CT981	Single use	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	14/09/2022	14/09/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	14/09/2022	14/09/2022
					CT983	Surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	14/09/2022	14/09/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	14/09/2022	14/09/2022
7783	2022	All other reportable incidents	47265	Dual-chamber implantable pacemaker, rate-responsive	CT141	Electrodes/Leads and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	22/08/2022	22/08/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	22/08/2022	22/08/2022
					CT145	Cardiovascular devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	22/08/2022	22/08/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	22/08/2022	22/08/2022
					CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	22/08/2022	22/08/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	22/08/2022	22/08/2022
					CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	22/08/2022	22/08/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	22/08/2022	22/08/2022
					CT233	Surgical	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	22/08/2022	22/08/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	22/08/2022	22/08/2022
					CT2406	Implantable	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	22/08/2022	22/08/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	22/08/2022	22/08/2022
					CT2407	Active implantable	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	22/08/2022	22/08/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	22/08/2022	22/08/2022
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	22/08/2022	22/08/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	22/08/2022	22/08/2022
					CT2727	Active implant devices/ancillaries	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	22/08/2022	22/08/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	22/08/2022	22/08/2022
					CT2750	Implantable cardiac electrical stimulators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	22/08/2022	22/08/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	22/08/2022	22/08/2022
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	22/08/2022	22/08/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	22/08/2022	22/08/2022
					CT2985	Active	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	22/08/2022	22/08/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	22/08/2022	22/08/2022
					CT3001	Implantable metal	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	22/08/2022	22/08/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	22/08/2022	22/08/2022
					CT3002	Implantable synthetic polymer	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	22/08/2022	22/08/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	22/08/2022	22/08/2022
					CT3006	Implantable inorganic material	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	22/08/2022	22/08/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	22/08/2022	22/08/2022
					CT314	Electrically powered	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	22/08/2022	22/08/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	22/08/2022	22/08/2022
					CT321	Long-term surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	22/08/2022	22/08/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	22/08/2022	22/08/2022
					CT449	Generators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	22/08/2022	22/08/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	22/08/2022	22/08/2022
					CT981	Single use	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	22/08/2022	22/08/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	22/08/2022	22/08/2022
					CT983	Surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	22/08/2022	22/08/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	22/08/2022	22/08/2022
7784	2022	All other reportable incidents	35852	Single-chamber implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
7785	2022	All other reportable incidents	17882	Physiologic-monitoring defibrillation system	CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A07	Electrical /Electronic Property Problem		
							F27	Problem Identified During Non-Clinical Procedure	A07	Electrical /Electronic Property Problem		
7786	2022	All other reportable incidents	17882	Physiologic-monitoring defibrillation system	CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A0801	Failure to Calibrate		
							F27	Problem Identified During Non-Clinical Procedure	A0801	Failure to Calibrate		
7787	2022	All other reportable incidents	47265	Dual-chamber implantable pacemaker, rate-responsive	CT141	Electrodes/Leads and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070910	Undersensing		
							F26	No Health Consequences or Impact	A070910	Undersensing		
					CT145	Cardiovascular devices	E2403	No Clinical Signs, Symptoms or Conditions	A070910	Undersensing		
							F26	No Health Consequences or Impact	A070910	Undersensing		
					CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070910	Undersensing		
							F26	No Health Consequences or Impact	A070910	Undersensing		
					CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A070910	Undersensing		
							F26	No Health Consequences or Impact	A070910	Undersensing		
					CT233	Surgical	E2403	No Clinical Signs, Symptoms or Conditions	A070910	Undersensing		
							F26	No Health Consequences or Impact	A070910	Undersensing		
					CT2406	Implantable	E2403	No Clinical Signs, Symptoms or Conditions	A070910	Undersensing		
							F26	No Health Consequences or Impact	A070910	Undersensing		
					CT2407	Active implantable	E2403	No Clinical Signs, Symptoms or Conditions	A070910	Undersensing		
							F26	No Health Consequences or Impact	A070910	Undersensing		
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A070910	Undersensing		
							F26	No Health Consequences or Impact	A070910	Undersensing		
					CT2727	Active implant devices/ancillaries	E2403	No Clinical Signs, Symptoms or Conditions	A070910	Undersensing		
							F26	No Health Consequences or Impact	A070910	Undersensing		
					CT2750	Implantable cardiac electrical stimulators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070910	Undersensing		
							F26	No Health Consequences or Impact	A070910	Undersensing		
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A070910	Undersensing		
							F26	No Health Consequences or Impact	A070910	Undersensing		
					CT2985	Active	E2403	No Clinical Signs, Symptoms or Conditions	A070910	Undersensing		
							F26	No Health Consequences or Impact	A070910	Undersensing		
					CT3001	Implantable metal	E2403	No Clinical Signs, Symptoms or Conditions	A070910	Undersensing		
							F26	No Health Consequences or Impact	A070910	Undersensing		
					CT3002	Implantable synthetic polymer	E2403	No Clinical Signs, Symptoms or Conditions	A070910	Undersensing		
							F26	No Health Consequences or Impact	A070910	Undersensing		
					CT3006	Implantable inorganic material	E2403	No Clinical Signs, Symptoms or Conditions	A070910	Undersensing		
							F26	No Health Consequences or Impact	A070910	Undersensing		
					CT314	Electrically powered	E2403	No Clinical Signs, Symptoms or Conditions	A070910	Undersensing		
							F26	No Health Consequences or Impact	A070910	Undersensing		
					CT321	Long-term surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A070910	Undersensing		

						F26	No Health Consequences or Impact	A070910	Undersensing			
				CT449	Generators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070910	Undersensing			
						F26	No Health Consequences or Impact	A070910	Undersensing			
				CT981	Single use	E2403	No Clinical Signs, Symptoms or Conditions	A070910	Undersensing			
						F26	No Health Consequences or Impact	A070910	Undersensing			
				CT983	Surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A070910	Undersensing			
						F26	No Health Consequences or Impact	A070910	Undersensing			
7788	2022	All other reportable incidents	37265	Dual-chamber implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A071101	Difficult to Interrogate		
							F08	Hospitalization or Prolonged Hospitalization	A071101	Difficult to Interrogate		
							F23	Unexpected Medical Intervention	A071101	Difficult to Interrogate		
7789	2022	All other reportable incidents	35852	Single-chamber implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E2104	Electric Shock	A070909	Oversensing	13/09/2022	13/09/2022
									A071302	Inappropriate/Inadequate Shock/Stimulation	13/09/2022	13/09/2022
									A090801	Signal Artifact/Noise	13/09/2022	13/09/2022
							E2403	No Clinical Signs, Symptoms or Conditions	A070909	Oversensing	13/09/2022	13/09/2022
									A071302	Inappropriate/Inadequate Shock/Stimulation	13/09/2022	13/09/2022
									A090801	Signal Artifact/Noise	13/09/2022	13/09/2022
							F08	Hospitalization or Prolonged Hospitalization	A070909	Oversensing	13/09/2022	13/09/2022
									A071302	Inappropriate/Inadequate Shock/Stimulation	13/09/2022	13/09/2022
									A090801	Signal Artifact/Noise	13/09/2022	13/09/2022
							F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A070909	Oversensing	13/09/2022	13/09/2022
									A071302	Inappropriate/Inadequate Shock/Stimulation	13/09/2022	13/09/2022
									A090801	Signal Artifact/Noise	13/09/2022	13/09/2022
							F1905	Device Revision or Replacement	A070909	Oversensing	13/09/2022	13/09/2022
									A071302	Inappropriate/Inadequate Shock/Stimulation	13/09/2022	13/09/2022
									A090801	Signal Artifact/Noise	13/09/2022	13/09/2022
							F2203	Imaging Required	A070909	Oversensing	13/09/2022	13/09/2022
									A071302	Inappropriate/Inadequate Shock/Stimulation	13/09/2022	13/09/2022
									A090801	Signal Artifact/Noise	13/09/2022	13/09/2022
7790	2022	All other reportable incidents	17882	Physiologic-monitoring defibrillation system	CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A071301	Failure to Deliver Shock/Stimulation		
							F26	No Health Consequences or Impact	A071301	Failure to Deliver Shock/Stimulation		
7791	2022	All other reportable incidents	17882	Physiologic-monitoring defibrillation system	CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A150105	Key or Button Unresponsive/not Working		
							F27	Problem Identified During Non-Clinical Procedure	A150105	Key or Button Unresponsive/not Working		
7792	2022	All other reportable incidents	47265	Dual-chamber implantable pacemaker, rate-responsive	CT141	Electrodes/Leads and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery		
					CT145	Cardiovascular devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery		
					CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery		
					CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery		
					CT233	Surgical	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery		
					CT2406	Implantable	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery		
					CT2407	Active implantable	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery		
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery		
					CT2727	Active implant devices/ancillaries	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery		
					CT2750	Implantable cardiac electrical stimulators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery		
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery		
					CT2985	Active	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery		
					CT3001	Implantable metal	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery		
					CT3002	Implantable synthetic polymer	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery		
					CT3006	Implantable inorganic material	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery		
					CT314	Electrically powered	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery		
					CT321	Long-term surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery		
					CT449	Generators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery		
					CT981	Single use	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery		
					CT983	Surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery		
7793	2022	All other reportable incidents	47263	Cardiac resynchronization therapy implantable pacemaker	CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A0908	Incorrect, Inadequate or Imprecise Result or Readings	24/08/2022	24/08/2022
							F08	Hospitalization or Prolonged Hospitalization	A0908	Incorrect, Inadequate or Imprecise Result or Readings	24/08/2022	24/08/2022
							F1905	Device Revision or Replacement	A0908	Incorrect, Inadequate or Imprecise Result or Readings	24/08/2022	24/08/2022
					CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A0908	Incorrect, Inadequate or Imprecise Result or Readings	24/08/2022	24/08/2022
							F08	Hospitalization or Prolonged Hospitalization	A0908	Incorrect, Inadequate or Imprecise Result or Readings	24/08/2022	24/08/2022
							F1905	Device Revision or Replacement	A0908	Incorrect, Inadequate or Imprecise Result or Readings	24/08/2022	24/08/2022
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A0908	Incorrect, Inadequate or Imprecise Result or Readings	24/08/2022	24/08/2022
							F08	Hospitalization or Prolonged Hospitalization	A0908	Incorrect, Inadequate or Imprecise Result or Readings	24/08/2022	24/08/2022
							F1905	Device Revision or Replacement	A0908	Incorrect, Inadequate or Imprecise Result or Readings	24/08/2022	24/08/2022

					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A0908	Incorrect, Inadequate or Imprecise Result or Readings	24/08/2022	24/08/2022
							F08	Hospitalization or Prolonged Hospitalization	A0908	Incorrect, Inadequate or Imprecise Result or Readings	24/08/2022	24/08/2022
							F1905	Device Revision or Replacement	A0908	Incorrect, Inadequate or Imprecise Result or Readings	24/08/2022	24/08/2022
7794	2022	All other reportable incidents	35852	Single-chamber implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A072201	High Impedance		
							F26	No Health Consequences or Impact	A072201	High Impedance		
7795	2022	All other reportable incidents	35223	Endocardial/interventricular septal pacing lead	CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070909	Oversensing		
							F1901	Additional Surgery	A070909	Oversensing		
7796	2022	All other reportable incidents	35223	Endocardial/interventricular septal pacing lead	CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070909	Oversensing		
							F26	No Health Consequences or Impact	A070909	Oversensing		
7797	2022	Death	47270	Cardiac resynchronization therapy implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E060110	Ventricular Fibrillation	A0709	Device Sensing Problem		
									A070909	Oversensing		
									A070910	Undersensing		
									A071301	Failure to Deliver Shock/Stimulation		
							F02	Death	A0709	Device Sensing Problem		
									A070909	Oversensing		
									A070910	Undersensing		
									A071301	Failure to Deliver Shock/Stimulation		
7798	2022	All other reportable incidents	47265	Dual-chamber implantable pacemaker, rate-responsive	CT141	Electrodes/Leads and associated devices	E011903	Syncope/Fainting	A070909	Oversensing	09/03/2022	09/03/2022
									A0712	Pacing Problem	09/03/2022	09/03/2022
									A072201	High Impedance	09/03/2022	09/03/2022
									A090801	Signal Artifact/Noise	09/03/2022	09/03/2022
							E2403	No Clinical Signs, Symptoms or Conditions	A070909	Oversensing	09/03/2022	09/03/2022
									A0712	Pacing Problem	09/03/2022	09/03/2022
									A072201	High Impedance	09/03/2022	09/03/2022
									A090801	Signal Artifact/Noise	09/03/2022	09/03/2022
							F08	Hospitalization or Prolonged Hospitalization	A070909	Oversensing	09/03/2022	09/03/2022
									A0712	Pacing Problem	09/03/2022	09/03/2022
									A072201	High Impedance	09/03/2022	09/03/2022
									A090801	Signal Artifact/Noise	09/03/2022	09/03/2022
							F1001	Absence of Treatment	A070909	Oversensing	09/03/2022	09/03/2022
									A0712	Pacing Problem	09/03/2022	09/03/2022
									A072201	High Impedance	09/03/2022	09/03/2022
									A090801	Signal Artifact/Noise	09/03/2022	09/03/2022
							F1905	Device Revision or Replacement	A070909	Oversensing	09/03/2022	09/03/2022
									A0712	Pacing Problem	09/03/2022	09/03/2022
									A072201	High Impedance	09/03/2022	09/03/2022
									A090801	Signal Artifact/Noise	09/03/2022	09/03/2022
					CT145	Cardiovascular devices	E011903	Syncope/Fainting	A070909	Oversensing	09/03/2022	09/03/2022
									A0712	Pacing Problem	09/03/2022	09/03/2022
									A072201	High Impedance	09/03/2022	09/03/2022
									A090801	Signal Artifact/Noise	09/03/2022	09/03/2022
							E2403	No Clinical Signs, Symptoms or Conditions	A070909	Oversensing	09/03/2022	09/03/2022
									A0712	Pacing Problem	09/03/2022	09/03/2022
									A072201	High Impedance	09/03/2022	09/03/2022
									A090801	Signal Artifact/Noise	09/03/2022	09/03/2022
							F08	Hospitalization or Prolonged Hospitalization	A070909	Oversensing	09/03/2022	09/03/2022
									A0712	Pacing Problem	09/03/2022	09/03/2022
									A072201	High Impedance	09/03/2022	09/03/2022
									A090801	Signal Artifact/Noise	09/03/2022	09/03/2022
							F1001	Absence of Treatment	A070909	Oversensing	09/03/2022	09/03/2022
									A0712	Pacing Problem	09/03/2022	09/03/2022
									A072201	High Impedance	09/03/2022	09/03/2022
									A090801	Signal Artifact/Noise	09/03/2022	09/03/2022
							F1905	Device Revision or Replacement	A070909	Oversensing	09/03/2022	09/03/2022
									A0712	Pacing Problem	09/03/2022	09/03/2022
									A072201	High Impedance	09/03/2022	09/03/2022
									A090801	Signal Artifact/Noise	09/03/2022	09/03/2022
					CT1684	Cardiac pacemakers and associated devices	E011903	Syncope/Fainting	A070909	Oversensing	09/03/2022	09/03/2022
									A0712	Pacing Problem	09/03/2022	09/03/2022
									A072201	High Impedance	09/03/2022	09/03/2022
									A090801	Signal Artifact/Noise	09/03/2022	09/03/2022
							E2403	No Clinical Signs, Symptoms or Conditions	A070909	Oversensing	09/03/2022	09/03/2022
									A0712	Pacing Problem	09/03/2022	09/03/2022
									A072201	High Impedance	09/03/2022	09/03/2022
									A090801	Signal Artifact/Noise	09/03/2022	09/03/2022
							F08	Hospitalization or Prolonged Hospitalization	A070909	Oversensing	09/03/2022	09/03/2022
									A0712	Pacing Problem	09/03/2022	09/03/2022
									A072201	High Impedance	09/03/2022	09/03/2022
									A090801	Signal Artifact/Noise	09/03/2022	09/03/2022
							F1001	Absence of Treatment	A070909	Oversensing	09/03/2022	09/03/2022
									A0712	Pacing Problem	09/03/2022	09/03/2022
									A072201	High Impedance	09/03/2022	09/03/2022
									A090801	Signal Artifact/Noise	09/03/2022	09/03/2022
							F1905	Device Revision or Replacement	A070909	Oversensing	09/03/2022	09/03/2022
									A0712	Pacing Problem	09/03/2022	09/03/2022
									A072201	High Impedance	09/03/2022	09/03/2022
									A090801	Signal Artifact/Noise	09/03/2022	09/03/2022
					CT1685	Cardiac pacemakers	E011903	Syncope/Fainting	A070909	Oversensing	09/03/2022	09/03/2022
									A0712	Pacing Problem	09/03/2022	09/03/2022
									A072201	High Impedance	09/03/2022	09/03/2022
									A090801	Signal Artifact/Noise	09/03/2022	09/03/2022
							E2403	No Clinical Signs, Symptoms or Conditions	A070909	Oversensing	09/03/2022	09/03/2022
									A0712	Pacing Problem	09/03/2022	09/03/2022
									A072201	High Impedance	09/03/2022	09/03/2022
									A090801	Signal Artifact/Noise	09/03/2022	09/03/2022

							F08	Hospitalization or Prolonged Hospitalization	A070909	Oversensing	09/03/2022	09/03/2022
									A0712	Pacing Problem	09/03/2022	09/03/2022
									A072201	High Impedance	09/03/2022	09/03/2022
									A090801	Signal Artifact/Noise	09/03/2022	09/03/2022
							F1001	Absence of Treatment	A070909	Oversensing	09/03/2022	09/03/2022
									A0712	Pacing Problem	09/03/2022	09/03/2022
									A072201	High Impedance	09/03/2022	09/03/2022
									A090801	Signal Artifact/Noise	09/03/2022	09/03/2022
							F1905	Device Revision or Replacement	A070909	Oversensing	09/03/2022	09/03/2022
									A0712	Pacing Problem	09/03/2022	09/03/2022
									A072201	High Impedance	09/03/2022	09/03/2022
									A090801	Signal Artifact/Noise	09/03/2022	09/03/2022
					CT233	Surgical	E011903	Syncope/Fainting	A070909	Oversensing	09/03/2022	09/03/2022
									A0712	Pacing Problem	09/03/2022	09/03/2022
									A072201	High Impedance	09/03/2022	09/03/2022
									A090801	Signal Artifact/Noise	09/03/2022	09/03/2022
							E2403	No Clinical Signs, Symptoms or Conditions	A070909	Oversensing	09/03/2022	09/03/2022
									A0712	Pacing Problem	09/03/2022	09/03/2022
									A072201	High Impedance	09/03/2022	09/03/2022
									A090801	Signal Artifact/Noise	09/03/2022	09/03/2022
							F08	Hospitalization or Prolonged Hospitalization	A070909	Oversensing	09/03/2022	09/03/2022
									A0712	Pacing Problem	09/03/2022	09/03/2022
									A072201	High Impedance	09/03/2022	09/03/2022
									A090801	Signal Artifact/Noise	09/03/2022	09/03/2022
							F1001	Absence of Treatment	A070909	Oversensing	09/03/2022	09/03/2022
									A0712	Pacing Problem	09/03/2022	09/03/2022
									A072201	High Impedance	09/03/2022	09/03/2022
									A090801	Signal Artifact/Noise	09/03/2022	09/03/2022
							F1905	Device Revision or Replacement	A070909	Oversensing	09/03/2022	09/03/2022
									A0712	Pacing Problem	09/03/2022	09/03/2022
									A072201	High Impedance	09/03/2022	09/03/2022
									A090801	Signal Artifact/Noise	09/03/2022	09/03/2022
					CT2406	Implantable	E011903	Syncope/Fainting	A070909	Oversensing	09/03/2022	09/03/2022
									A0712	Pacing Problem	09/03/2022	09/03/2022
									A072201	High Impedance	09/03/2022	09/03/2022
									A090801	Signal Artifact/Noise	09/03/2022	09/03/2022
							E2403	No Clinical Signs, Symptoms or Conditions	A070909	Oversensing	09/03/2022	09/03/2022
									A0712	Pacing Problem	09/03/2022	09/03/2022
									A072201	High Impedance	09/03/2022	09/03/2022
									A090801	Signal Artifact/Noise	09/03/2022	09/03/2022
							F08	Hospitalization or Prolonged Hospitalization	A070909	Oversensing	09/03/2022	09/03/2022
									A0712	Pacing Problem	09/03/2022	09/03/2022
									A072201	High Impedance	09/03/2022	09/03/2022
									A090801	Signal Artifact/Noise	09/03/2022	09/03/2022
							F1001	Absence of Treatment	A070909	Oversensing	09/03/2022	09/03/2022
									A0712	Pacing Problem	09/03/2022	09/03/2022
									A072201	High Impedance	09/03/2022	09/03/2022
									A090801	Signal Artifact/Noise	09/03/2022	09/03/2022
							F1905	Device Revision or Replacement	A070909	Oversensing	09/03/2022	09/03/2022
									A0712	Pacing Problem	09/03/2022	09/03/2022
									A072201	High Impedance	09/03/2022	09/03/2022
									A090801	Signal Artifact/Noise	09/03/2022	09/03/2022
					CT2407	Active implantable	E011903	Syncope/Fainting	A070909	Oversensing	09/03/2022	09/03/2022
									A0712	Pacing Problem	09/03/2022	09/03/2022
									A072201	High Impedance	09/03/2022	09/03/2022
									A090801	Signal Artifact/Noise	09/03/2022	09/03/2022
							E2403	No Clinical Signs, Symptoms or Conditions	A070909	Oversensing	09/03/2022	09/03/2022
									A0712	Pacing Problem	09/03/2022	09/03/2022
									A072201	High Impedance	09/03/2022	09/03/2022
									A090801	Signal Artifact/Noise	09/03/2022	09/03/2022
							F08	Hospitalization or Prolonged Hospitalization	A070909	Oversensing	09/03/2022	09/03/2022
									A0712	Pacing Problem	09/03/2022	09/03/2022
									A072201	High Impedance	09/03/2022	09/03/2022
									A090801	Signal Artifact/Noise	09/03/2022	09/03/2022
							F1001	Absence of Treatment	A070909	Oversensing	09/03/2022	09/03/2022
									A0712	Pacing Problem	09/03/2022	09/03/2022
									A072201	High Impedance	09/03/2022	09/03/2022
									A090801	Signal Artifact/Noise	09/03/2022	09/03/2022
							F1905	Device Revision or Replacement	A070909	Oversensing	09/03/2022	09/03/2022
									A0712	Pacing Problem	09/03/2022	09/03/2022
									A072201	High Impedance	09/03/2022	09/03/2022
									A090801	Signal Artifact/Noise	09/03/2022	09/03/2022
					CT257	Implantable cardiac pacemakers	E011903	Syncope/Fainting	A070909	Oversensing	09/03/2022	09/03/2022
									A0712	Pacing Problem	09/03/2022	09/03/2022
									A072201	High Impedance	09/03/2022	09/03/2022
									A090801	Signal Artifact/Noise	09/03/2022	09/03/2022
							E2403	No Clinical Signs, Symptoms or Conditions	A070909	Oversensing	09/03/2022	09/03/2022
									A0712	Pacing Problem	09/03/2022	09/03/2022
									A072201	High Impedance	09/03/2022	09/03/2022
									A090801	Signal Artifact/Noise	09/03/2022	09/03/2022
							F08	Hospitalization or Prolonged Hospitalization	A070909	Oversensing	09/03/2022	09/03/2022
									A0712	Pacing Problem	09/03/2022	09/03/2022
									A072201	High Impedance	09/03/2022	09/03/2022
									A090801	Signal Artifact/Noise	09/03/2022	09/03/2022
							F1001	Absence of Treatment	A070909	Oversensing	09/03/2022	09/03/2022
									A0712	Pacing Problem	09/03/2022	09/03/2022
									A072201	High Impedance	09/03/2022	09/03/2022
									A090801	Signal Artifact/Noise	09/03/2022	09/03/2022

							F1905	Device Revision or Replacement	A070909	Oversensing	09/03/2022	09/03/2022
									A0712	Pacing Problem	09/03/2022	09/03/2022
									A072201	High Impedance	09/03/2022	09/03/2022
									A090801	Signal Artifact/Noise	09/03/2022	09/03/2022
					CT2727	Active implant devices/ancillaries	E011903	Syncope/Fainting	A070909	Oversensing	09/03/2022	09/03/2022
									A0712	Pacing Problem	09/03/2022	09/03/2022
									A072201	High Impedance	09/03/2022	09/03/2022
									A090801	Signal Artifact/Noise	09/03/2022	09/03/2022
							E2403	No Clinical Signs, Symptoms or Conditions	A070909	Oversensing	09/03/2022	09/03/2022
									A0712	Pacing Problem	09/03/2022	09/03/2022
									A072201	High Impedance	09/03/2022	09/03/2022
									A090801	Signal Artifact/Noise	09/03/2022	09/03/2022
							F08	Hospitalization or Prolonged Hospitalization	A070909	Oversensing	09/03/2022	09/03/2022
									A0712	Pacing Problem	09/03/2022	09/03/2022
									A072201	High Impedance	09/03/2022	09/03/2022
									A090801	Signal Artifact/Noise	09/03/2022	09/03/2022
							F1001	Absence of Treatment	A070909	Oversensing	09/03/2022	09/03/2022
									A0712	Pacing Problem	09/03/2022	09/03/2022
									A072201	High Impedance	09/03/2022	09/03/2022
									A090801	Signal Artifact/Noise	09/03/2022	09/03/2022
							F1905	Device Revision or Replacement	A070909	Oversensing	09/03/2022	09/03/2022
									A0712	Pacing Problem	09/03/2022	09/03/2022
									A072201	High Impedance	09/03/2022	09/03/2022
									A090801	Signal Artifact/Noise	09/03/2022	09/03/2022
					CT2750	Implantable cardiac electrical stimulators and associated devices	E011903	Syncope/Fainting	A070909	Oversensing	09/03/2022	09/03/2022
									A0712	Pacing Problem	09/03/2022	09/03/2022
									A072201	High Impedance	09/03/2022	09/03/2022
									A090801	Signal Artifact/Noise	09/03/2022	09/03/2022
							E2403	No Clinical Signs, Symptoms or Conditions	A070909	Oversensing	09/03/2022	09/03/2022
									A0712	Pacing Problem	09/03/2022	09/03/2022
									A072201	High Impedance	09/03/2022	09/03/2022
									A090801	Signal Artifact/Noise	09/03/2022	09/03/2022
							F08	Hospitalization or Prolonged Hospitalization	A070909	Oversensing	09/03/2022	09/03/2022
									A0712	Pacing Problem	09/03/2022	09/03/2022
									A072201	High Impedance	09/03/2022	09/03/2022
									A090801	Signal Artifact/Noise	09/03/2022	09/03/2022
							F1001	Absence of Treatment	A070909	Oversensing	09/03/2022	09/03/2022
									A0712	Pacing Problem	09/03/2022	09/03/2022
									A072201	High Impedance	09/03/2022	09/03/2022
									A090801	Signal Artifact/Noise	09/03/2022	09/03/2022
							F1905	Device Revision or Replacement	A070909	Oversensing	09/03/2022	09/03/2022
									A0712	Pacing Problem	09/03/2022	09/03/2022
									A072201	High Impedance	09/03/2022	09/03/2022
									A090801	Signal Artifact/Noise	09/03/2022	09/03/2022
					CT2866	Implantable cardiac electrical stimulators	E011903	Syncope/Fainting	A070909	Oversensing	09/03/2022	09/03/2022
									A0712	Pacing Problem	09/03/2022	09/03/2022
									A072201	High Impedance	09/03/2022	09/03/2022
									A090801	Signal Artifact/Noise	09/03/2022	09/03/2022
							E2403	No Clinical Signs, Symptoms or Conditions	A070909	Oversensing	09/03/2022	09/03/2022
									A0712	Pacing Problem	09/03/2022	09/03/2022
									A072201	High Impedance	09/03/2022	09/03/2022
									A090801	Signal Artifact/Noise	09/03/2022	09/03/2022
							F08	Hospitalization or Prolonged Hospitalization	A070909	Oversensing	09/03/2022	09/03/2022
									A0712	Pacing Problem	09/03/2022	09/03/2022
									A072201	High Impedance	09/03/2022	09/03/2022
									A090801	Signal Artifact/Noise	09/03/2022	09/03/2022
							F1001	Absence of Treatment	A070909	Oversensing	09/03/2022	09/03/2022
									A0712	Pacing Problem	09/03/2022	09/03/2022
									A072201	High Impedance	09/03/2022	09/03/2022
									A090801	Signal Artifact/Noise	09/03/2022	09/03/2022
							F1905	Device Revision or Replacement	A070909	Oversensing	09/03/2022	09/03/2022
									A0712	Pacing Problem	09/03/2022	09/03/2022
									A072201	High Impedance	09/03/2022	09/03/2022
									A090801	Signal Artifact/Noise	09/03/2022	09/03/2022
					CT2985	Active	E011903	Syncope/Fainting	A070909	Oversensing	09/03/2022	09/03/2022
									A0712	Pacing Problem	09/03/2022	09/03/2022
									A072201	High Impedance	09/03/2022	09/03/2022
									A090801	Signal Artifact/Noise	09/03/2022	09/03/2022
							E2403	No Clinical Signs, Symptoms or Conditions	A070909	Oversensing	09/03/2022	09/03/2022
									A0712	Pacing Problem	09/03/2022	09/03/2022
									A072201	High Impedance	09/03/2022	09/03/2022
									A090801	Signal Artifact/Noise	09/03/2022	09/03/2022
							F08	Hospitalization or Prolonged Hospitalization	A070909	Oversensing	09/03/2022	09/03/2022
									A0712	Pacing Problem	09/03/2022	09/03/2022
									A072201	High Impedance	09/03/2022	09/03/2022
									A090801	Signal Artifact/Noise	09/03/2022	09/03/2022
							F1001	Absence of Treatment	A070909	Oversensing	09/03/2022	09/03/2022
									A0712	Pacing Problem	09/03/2022	09/03/2022
									A072201	High Impedance	09/03/2022	09/03/2022
									A090801	Signal Artifact/Noise	09/03/2022	09/03/2022
							F1905	Device Revision or Replacement	A070909	Oversensing	09/03/2022	09/03/2022
									A0712	Pacing Problem	09/03/2022	09/03/2022
									A072201	High Impedance	09/03/2022	09/03/2022
									A090801	Signal Artifact/Noise	09/03/2022	09/03/2022
					CT3001	Implantable metal	E011903	Syncope/Fainting	A070909	Oversensing	09/03/2022	09/03/2022
									A0712	Pacing Problem	09/03/2022	09/03/2022
									A072201	High Impedance	09/03/2022	09/03/2022
									A090801	Signal Artifact/Noise	09/03/2022	09/03/2022

							E2403	No Clinical Signs, Symptoms or Conditions	A070909	Oversensing	09/03/2022	09/03/2022
									A0712	Pacing Problem	09/03/2022	09/03/2022
									A072201	High Impedance	09/03/2022	09/03/2022
									A090801	Signal Artifact/Noise	09/03/2022	09/03/2022
							F08	Hospitalization or Prolonged Hospitalization	A070909	Oversensing	09/03/2022	09/03/2022
									A0712	Pacing Problem	09/03/2022	09/03/2022
									A072201	High Impedance	09/03/2022	09/03/2022
									A090801	Signal Artifact/Noise	09/03/2022	09/03/2022
							F1001	Absence of Treatment	A070909	Oversensing	09/03/2022	09/03/2022
									A0712	Pacing Problem	09/03/2022	09/03/2022
									A072201	High Impedance	09/03/2022	09/03/2022
									A090801	Signal Artifact/Noise	09/03/2022	09/03/2022
							F1905	Device Revision or Replacement	A070909	Oversensing	09/03/2022	09/03/2022
									A0712	Pacing Problem	09/03/2022	09/03/2022
									A072201	High Impedance	09/03/2022	09/03/2022
									A090801	Signal Artifact/Noise	09/03/2022	09/03/2022
						CT3002	Implantable synthetic polymer	E011903	Syncope/Fainting	A070909	Oversensing	09/03/2022	09/03/2022
										A0712	Pacing Problem	09/03/2022	09/03/2022
										A072201	High Impedance	09/03/2022	09/03/2022
										A090801	Signal Artifact/Noise	09/03/2022	09/03/2022
								E2403	No Clinical Signs, Symptoms or Conditions	A070909	Oversensing	09/03/2022	09/03/2022
										A0712	Pacing Problem	09/03/2022	09/03/2022
										A072201	High Impedance	09/03/2022	09/03/2022
										A090801	Signal Artifact/Noise	09/03/2022	09/03/2022
								F08	Hospitalization or Prolonged Hospitalization	A070909	Oversensing	09/03/2022	09/03/2022
										A0712	Pacing Problem	09/03/2022	09/03/2022
										A072201	High Impedance	09/03/2022	09/03/2022
										A090801	Signal Artifact/Noise	09/03/2022	09/03/2022
								F1001	Absence of Treatment	A070909	Oversensing	09/03/2022	09/03/2022
										A0712	Pacing Problem	09/03/2022	09/03/2022
										A072201	High Impedance	09/03/2022	09/03/2022
										A090801	Signal Artifact/Noise	09/03/2022	09/03/2022
								F1905	Device Revision or Replacement	A070909	Oversensing	09/03/2022	09/03/2022
										A0712	Pacing Problem	09/03/2022	09/03/2022
										A072201	High Impedance	09/03/2022	09/03/2022
										A090801	Signal Artifact/Noise	09/03/2022	09/03/2022
						CT3006	Implantable inorganic material	E011903	Syncope/Fainting	A070909	Oversensing	09/03/2022	09/03/2022
										A0712	Pacing Problem	09/03/2022	09/03/2022
										A072201	High Impedance	09/03/2022	09/03/2022
										A090801	Signal Artifact/Noise	09/03/2022	09/03/2022
								E2403	No Clinical Signs, Symptoms or Conditions	A070909	Oversensing	09/03/2022	09/03/2022
										A0712	Pacing Problem	09/03/2022	09/03/2022
										A072201	High Impedance	09/03/2022	09/03/2022
										A090801	Signal Artifact/Noise	09/03/2022	09/03/2022
								F08	Hospitalization or Prolonged Hospitalization	A070909	Oversensing	09/03/2022	09/03/2022
										A0712	Pacing Problem	09/03/2022	09/03/2022
										A072201	High Impedance	09/03/2022	09/03/2022
										A090801	Signal Artifact/Noise	09/03/2022	09/03/2022
								F1001	Absence of Treatment	A070909	Oversensing	09/03/2022	09/03/2022
										A0712	Pacing Problem	09/03/2022	09/03/2022
										A072201	High Impedance	09/03/2022	09/03/2022
										A090801	Signal Artifact/Noise	09/03/2022	09/03/2022
								F1905	Device Revision or Replacement	A070909	Oversensing	09/03/2022	09/03/2022
										A0712	Pacing Problem	09/03/2022	09/03/2022
										A072201	High Impedance	09/03/2022	09/03/2022
										A090801	Signal Artifact/Noise	09/03/2022	09/03/2022
						CT314	Electrically powered	E011903	Syncope/Fainting	A070909	Oversensing	09/03/2022	09/03/2022
										A0712	Pacing Problem	09/03/2022	09/03/2022
										A072201	High Impedance	09/03/2022	09/03/2022
										A090801	Signal Artifact/Noise	09/03/2022	09/03/2022
								E2403	No Clinical Signs, Symptoms or Conditions	A070909	Oversensing	09/03/2022	09/03/2022
										A0712	Pacing Problem	09/03/2022	09/03/2022
										A072201	High Impedance	09/03/2022	09/03/2022
										A090801	Signal Artifact/Noise	09/03/2022	09/03/2022
								F08	Hospitalization or Prolonged Hospitalization	A070909	Oversensing	09/03/2022	09/03/2022
										A0712	Pacing Problem	09/03/2022	09/03/2022
										A072201	High Impedance	09/03/2022	09/03/2022
										A090801	Signal Artifact/Noise	09/03/2022	09/03/2022
								F1001	Absence of Treatment	A070909	Oversensing	09/03/2022	09/03/2022
										A0712	Pacing Problem	09/03/2022	09/03/2022
										A072201	High Impedance	09/03/2022	09/03/2022
										A090801	Signal Artifact/Noise	09/03/2022	09/03/2022
								F1905	Device Revision or Replacement	A070909	Oversensing	09/03/2022	09/03/2022
										A0712	Pacing Problem	09/03/2022	09/03/2022
										A072201	High Impedance	09/03/2022	09/03/2022
										A090801	Signal Artifact/Noise	09/03/2022	09/03/2022
						CT321	Long-term surgical invasive	E011903	Syncope/Fainting	A070909	Oversensing	09/03/2022	09/03/2022
										A0712	Pacing Problem	09/03/2022	09/03/2022
										A072201	High Impedance	09/03/2022	09/03/2022
										A090801	Signal Artifact/Noise	09/03/2022	09/03/2022
								E2403	No Clinical Signs, Symptoms or Conditions	A070909	Oversensing	09/03/2022	09/03/2022
										A0712	Pacing Problem	09/03/2022	09/03/2022
										A072201	High Impedance	09/03/2022	09/03/2022
										A090801	Signal Artifact/Noise	09/03/2022	09/03/2022
								F08	Hospitalization or Prolonged Hospitalization	A070909	Oversensing	09/03/2022	09/03/2022
										A0712	Pacing Problem	09/03/2022	09/03/2022
										A072201	High Impedance	09/03/2022	09/03/2022
										A090801	Signal Artifact/Noise	09/03/2022	09/03/2022

							F1001	Absence of Treatment	A070909	Oversensing	09/03/2022	09/03/2022
									A0712	Pacing Problem	09/03/2022	09/03/2022
									A072201	High Impedance	09/03/2022	09/03/2022
									A090801	Signal Artifact/Noise	09/03/2022	09/03/2022
							F1905	Device Revision or Replacement	A070909	Oversensing	09/03/2022	09/03/2022
									A0712	Pacing Problem	09/03/2022	09/03/2022
									A072201	High Impedance	09/03/2022	09/03/2022
									A090801	Signal Artifact/Noise	09/03/2022	09/03/2022
					CT449	Generators and associated devices	E011903	Syncope/Fainting	A070909	Oversensing	09/03/2022	09/03/2022
									A0712	Pacing Problem	09/03/2022	09/03/2022
									A072201	High Impedance	09/03/2022	09/03/2022
									A090801	Signal Artifact/Noise	09/03/2022	09/03/2022
							E2403	No Clinical Signs, Symptoms or Conditions	A070909	Oversensing	09/03/2022	09/03/2022
									A0712	Pacing Problem	09/03/2022	09/03/2022
									A072201	High Impedance	09/03/2022	09/03/2022
									A090801	Signal Artifact/Noise	09/03/2022	09/03/2022
							F08	Hospitalization or Prolonged Hospitalization	A070909	Oversensing	09/03/2022	09/03/2022
									A0712	Pacing Problem	09/03/2022	09/03/2022
									A072201	High Impedance	09/03/2022	09/03/2022
									A090801	Signal Artifact/Noise	09/03/2022	09/03/2022
							F1001	Absence of Treatment	A070909	Oversensing	09/03/2022	09/03/2022
									A0712	Pacing Problem	09/03/2022	09/03/2022
									A072201	High Impedance	09/03/2022	09/03/2022
									A090801	Signal Artifact/Noise	09/03/2022	09/03/2022
							F1905	Device Revision or Replacement	A070909	Oversensing	09/03/2022	09/03/2022
									A0712	Pacing Problem	09/03/2022	09/03/2022
									A072201	High Impedance	09/03/2022	09/03/2022
									A090801	Signal Artifact/Noise	09/03/2022	09/03/2022
					CT981	Single use	E011903	Syncope/Fainting	A070909	Oversensing	09/03/2022	09/03/2022
									A0712	Pacing Problem	09/03/2022	09/03/2022
									A072201	High Impedance	09/03/2022	09/03/2022
									A090801	Signal Artifact/Noise	09/03/2022	09/03/2022
							E2403	No Clinical Signs, Symptoms or Conditions	A070909	Oversensing	09/03/2022	09/03/2022
									A0712	Pacing Problem	09/03/2022	09/03/2022
									A072201	High Impedance	09/03/2022	09/03/2022
									A090801	Signal Artifact/Noise	09/03/2022	09/03/2022
							F08	Hospitalization or Prolonged Hospitalization	A070909	Oversensing	09/03/2022	09/03/2022
									A0712	Pacing Problem	09/03/2022	09/03/2022
									A072201	High Impedance	09/03/2022	09/03/2022
									A090801	Signal Artifact/Noise	09/03/2022	09/03/2022
							F1001	Absence of Treatment	A070909	Oversensing	09/03/2022	09/03/2022
									A0712	Pacing Problem	09/03/2022	09/03/2022
									A072201	High Impedance	09/03/2022	09/03/2022
									A090801	Signal Artifact/Noise	09/03/2022	09/03/2022
							F1905	Device Revision or Replacement	A070909	Oversensing	09/03/2022	09/03/2022
									A0712	Pacing Problem	09/03/2022	09/03/2022
									A072201	High Impedance	09/03/2022	09/03/2022
									A090801	Signal Artifact/Noise	09/03/2022	09/03/2022
					CT983	Surgical invasive	E011903	Syncope/Fainting	A070909	Oversensing	09/03/2022	09/03/2022
									A0712	Pacing Problem	09/03/2022	09/03/2022
									A072201	High Impedance	09/03/2022	09/03/2022
									A090801	Signal Artifact/Noise	09/03/2022	09/03/2022
							E2403	No Clinical Signs, Symptoms or Conditions	A070909	Oversensing	09/03/2022	09/03/2022
									A0712	Pacing Problem	09/03/2022	09/03/2022
									A072201	High Impedance	09/03/2022	09/03/2022
									A090801	Signal Artifact/Noise	09/03/2022	09/03/2022
							F08	Hospitalization or Prolonged Hospitalization	A070909	Oversensing	09/03/2022	09/03/2022
									A0712	Pacing Problem	09/03/2022	09/03/2022
									A072201	High Impedance	09/03/2022	09/03/2022
									A090801	Signal Artifact/Noise	09/03/2022	09/03/2022
							F1001	Absence of Treatment	A070909	Oversensing	09/03/2022	09/03/2022
									A0712	Pacing Problem	09/03/2022	09/03/2022
									A072201	High Impedance	09/03/2022	09/03/2022
									A090801	Signal Artifact/Noise	09/03/2022	09/03/2022
							F1905	Device Revision or Replacement	A070909	Oversensing	09/03/2022	09/03/2022
									A0712	Pacing Problem	09/03/2022	09/03/2022
									A072201	High Impedance	09/03/2022	09/03/2022
									A090801	Signal Artifact/Noise	09/03/2022	09/03/2022
7799	2022	All other reportable incidents	36241	Pacemaker/defibrillator lead	CT1684	Cardiac pacemakers and associated devices	E011903	Syncope/Fainting	A040101	Fracture		
									A070909	Oversensing		
									A0712	Pacing Problem		
									A072201	High Impedance		
									A090801	Signal Artifact/Noise		
							E2403	No Clinical Signs, Symptoms or Conditions	A040101	Fracture		
									A070909	Oversensing		
									A0712	Pacing Problem		
									A072201	High Impedance		
									A090801	Signal Artifact/Noise		
							F1001	Absence of Treatment	A040101	Fracture		
									A070909	Oversensing		
									A0712	Pacing Problem		
									A072201	High Impedance		
									A090801	Signal Artifact/Noise		
7800	2022	All other reportable incidents	35223	Endocardial/interventricular septal pacing lead	CT1684	Cardiac pacemakers and associated devices	E0112	Dizziness	A040101	Fracture	03/02/2022	03/02/2022
							E060106	Heart Block	A040101	Fracture	03/02/2022	03/02/2022
							E2403	No Clinical Signs, Symptoms or Conditions	A040101	Fracture	03/02/2022	03/02/2022
							F08	Hospitalization or Prolonged Hospitalization	A040101	Fracture	03/02/2022	03/02/2022
							F11	Non-Serious Injury/ Illness/ Impairment	A040101	Fracture	03/02/2022	03/02/2022

							F12	Serious Injury/ Illness/ Impairment	A040101	Fracture	03/02/2022	03/02/2022
							F1903	Device Explantation	A040101	Fracture	03/02/2022	03/02/2022
7801	2022	All other reportable incidents	35223	Endocardial/interventricular septal pacing lead	CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070909	Oversensing		
							F26	No Health Consequences or Impact	A070909	Oversensing		
7802	2022	All other reportable incidents	47265	Dual-chamber implantable pacemaker, rate-responsive	CT141	Electrodes/Leads and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	18/09/2022	18/09/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	18/09/2022	18/09/2022
					CT145	Cardiovascular devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	18/09/2022	18/09/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	18/09/2022	18/09/2022
					CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	18/09/2022	18/09/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	18/09/2022	18/09/2022
					CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	18/09/2022	18/09/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	18/09/2022	18/09/2022
					CT233	Surgical	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	18/09/2022	18/09/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	18/09/2022	18/09/2022
					CT2406	Implantable	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	18/09/2022	18/09/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	18/09/2022	18/09/2022
					CT2407	Active implantable	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	18/09/2022	18/09/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	18/09/2022	18/09/2022
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	18/09/2022	18/09/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	18/09/2022	18/09/2022
					CT2727	Active implant devices/ancillaries	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	18/09/2022	18/09/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	18/09/2022	18/09/2022
					CT2750	Implantable cardiac electrical stimulators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	18/09/2022	18/09/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	18/09/2022	18/09/2022
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	18/09/2022	18/09/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	18/09/2022	18/09/2022
					CT2985	Active	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	18/09/2022	18/09/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	18/09/2022	18/09/2022
					CT3001	Implantable metal	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	18/09/2022	18/09/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	18/09/2022	18/09/2022
					CT3002	Implantable synthetic polymer	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	18/09/2022	18/09/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	18/09/2022	18/09/2022
					CT3006	Implantable inorganic material	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	18/09/2022	18/09/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	18/09/2022	18/09/2022
					CT314	Electrically powered	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	18/09/2022	18/09/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	18/09/2022	18/09/2022
					CT321	Long-term surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	18/09/2022	18/09/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	18/09/2022	18/09/2022
					CT449	Generators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	18/09/2022	18/09/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	18/09/2022	18/09/2022
					CT981	Single use	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	18/09/2022	18/09/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	18/09/2022	18/09/2022
					CT983	Surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	18/09/2022	18/09/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	18/09/2022	18/09/2022
7803	2022		47265	Dual-chamber implantable pacemaker, rate-responsive	CT141	Electrodes/Leads and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	18/09/2022	18/09/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	18/09/2022	18/09/2022
					CT145	Cardiovascular devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	18/09/2022	18/09/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	18/09/2022	18/09/2022
					CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	18/09/2022	18/09/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	18/09/2022	18/09/2022
					CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	18/09/2022	18/09/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	18/09/2022	18/09/2022
					CT233	Surgical	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	18/09/2022	18/09/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	18/09/2022	18/09/2022
					CT2406	Implantable	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	18/09/2022	18/09/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	18/09/2022	18/09/2022
					CT2407	Active implantable	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	18/09/2022	18/09/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	18/09/2022	18/09/2022
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	18/09/2022	18/09/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	18/09/2022	18/09/2022
					CT2727	Active implant devices/ancillaries	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	18/09/2022	18/09/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	18/09/2022	18/09/2022
					CT2750	Implantable cardiac electrical stimulators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	18/09/2022	18/09/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	18/09/2022	18/09/2022
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	18/09/2022	18/09/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	18/09/2022	18/09/2022
					CT2985	Active	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	18/09/2022	18/09/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	18/09/2022	18/09/2022
					CT3001	Implantable metal	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	18/09/2022	18/09/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	18/09/2022	18/09/2022
					CT3002	Implantable synthetic polymer	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	18/09/2022	18/09/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	18/09/2022	18/09/2022
					CT3006	Implantable inorganic material	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	18/09/2022	18/09/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	18/09/2022	18/09/2022
					CT314	Electrically powered	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	18/09/2022	18/09/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	18/09/2022	18/09/2022
					CT321	Long-term surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	18/09/2022	18/09/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	18/09/2022	18/09/2022
					CT449	Generators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	18/09/2022	18/09/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	18/09/2022	18/09/2022
					CT981	Single use	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	18/09/2022	18/09/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	18/09/2022	18/09/2022
					CT983	Surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	18/09/2022	18/09/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	18/09/2022	18/09/2022
7804	2022	All other reportable incidents	17882	Physiologic-monitoring defibrillation system	CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A071302	Inappropriate/Inadequate Shock/Stimulation		
							F27	Problem Identified During Non-Clinical Procedure	A071302	Inappropriate/Inadequate Shock/Stimulation		
7805	2022	All other reportable incidents	17882	Physiologic-monitoring defibrillation system	CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070803	Failure to Power Up		
							F27	Problem Identified During Non-Clinical Procedure	A070803	Failure to Power Up		

7806	2022	All other reportable incidents	12913	Pacemaker, implantable	CT145	Cardiovascular devices	E2104	Electric Shock	A070909	Oversensing	13/09/2022	13/09/2022
									A071302	Inappropriate/Inadequate Shock/Stimulation	13/09/2022	13/09/2022
									A090801	Signal Artifact/Noise	13/09/2022	13/09/2022
							E2403	No Clinical Signs, Symptoms or Conditions	A070909	Oversensing	13/09/2022	13/09/2022
									A071302	Inappropriate/Inadequate Shock/Stimulation	13/09/2022	13/09/2022
									A090801	Signal Artifact/Noise	13/09/2022	13/09/2022
							F08	Hospitalization or Prolonged Hospitalization	A070909	Oversensing	13/09/2022	13/09/2022
									A071302	Inappropriate/Inadequate Shock/Stimulation	13/09/2022	13/09/2022
									A090801	Signal Artifact/Noise	13/09/2022	13/09/2022
							F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A070909	Oversensing	13/09/2022	13/09/2022
									A071302	Inappropriate/Inadequate Shock/Stimulation	13/09/2022	13/09/2022
									A090801	Signal Artifact/Noise	13/09/2022	13/09/2022
							F1905	Device Revision or Replacement	A070909	Oversensing	13/09/2022	13/09/2022
									A071302	Inappropriate/Inadequate Shock/Stimulation	13/09/2022	13/09/2022
									A090801	Signal Artifact/Noise	13/09/2022	13/09/2022
							F2203	Imaging Required	A070909	Oversensing	13/09/2022	13/09/2022
									A071302	Inappropriate/Inadequate Shock/Stimulation	13/09/2022	13/09/2022
									A090801	Signal Artifact/Noise	13/09/2022	13/09/2022
					CT1684	Cardiac pacemakers and associated devices	E2104	Electric Shock	A070909	Oversensing	13/09/2022	13/09/2022
									A071302	Inappropriate/Inadequate Shock/Stimulation	13/09/2022	13/09/2022
									A090801	Signal Artifact/Noise	13/09/2022	13/09/2022
							E2403	No Clinical Signs, Symptoms or Conditions	A070909	Oversensing	13/09/2022	13/09/2022
									A071302	Inappropriate/Inadequate Shock/Stimulation	13/09/2022	13/09/2022
									A090801	Signal Artifact/Noise	13/09/2022	13/09/2022
							F08	Hospitalization or Prolonged Hospitalization	A070909	Oversensing	13/09/2022	13/09/2022
									A071302	Inappropriate/Inadequate Shock/Stimulation	13/09/2022	13/09/2022
									A090801	Signal Artifact/Noise	13/09/2022	13/09/2022
							F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A070909	Oversensing	13/09/2022	13/09/2022
									A071302	Inappropriate/Inadequate Shock/Stimulation	13/09/2022	13/09/2022
									A090801	Signal Artifact/Noise	13/09/2022	13/09/2022
							F1905	Device Revision or Replacement	A070909	Oversensing	13/09/2022	13/09/2022
									A071302	Inappropriate/Inadequate Shock/Stimulation	13/09/2022	13/09/2022
									A090801	Signal Artifact/Noise	13/09/2022	13/09/2022
							F2203	Imaging Required	A070909	Oversensing	13/09/2022	13/09/2022
									A071302	Inappropriate/Inadequate Shock/Stimulation	13/09/2022	13/09/2022
									A090801	Signal Artifact/Noise	13/09/2022	13/09/2022
					CT1685	Cardiac pacemakers	E2104	Electric Shock	A070909	Oversensing	13/09/2022	13/09/2022
									A071302	Inappropriate/Inadequate Shock/Stimulation	13/09/2022	13/09/2022
									A090801	Signal Artifact/Noise	13/09/2022	13/09/2022
							E2403	No Clinical Signs, Symptoms or Conditions	A070909	Oversensing	13/09/2022	13/09/2022
									A071302	Inappropriate/Inadequate Shock/Stimulation	13/09/2022	13/09/2022
									A090801	Signal Artifact/Noise	13/09/2022	13/09/2022
							F08	Hospitalization or Prolonged Hospitalization	A070909	Oversensing	13/09/2022	13/09/2022
									A071302	Inappropriate/Inadequate Shock/Stimulation	13/09/2022	13/09/2022
									A090801	Signal Artifact/Noise	13/09/2022	13/09/2022
							F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A070909	Oversensing	13/09/2022	13/09/2022
									A071302	Inappropriate/Inadequate Shock/Stimulation	13/09/2022	13/09/2022
									A090801	Signal Artifact/Noise	13/09/2022	13/09/2022
							F1905	Device Revision or Replacement	A070909	Oversensing	13/09/2022	13/09/2022
									A071302	Inappropriate/Inadequate Shock/Stimulation	13/09/2022	13/09/2022
									A090801	Signal Artifact/Noise	13/09/2022	13/09/2022
							F2203	Imaging Required	A070909	Oversensing	13/09/2022	13/09/2022
									A071302	Inappropriate/Inadequate Shock/Stimulation	13/09/2022	13/09/2022
									A090801	Signal Artifact/Noise	13/09/2022	13/09/2022
					CT2406	Implantable	E2104	Electric Shock	A070909	Oversensing	13/09/2022	13/09/2022
									A071302	Inappropriate/Inadequate Shock/Stimulation	13/09/2022	13/09/2022
									A090801	Signal Artifact/Noise	13/09/2022	13/09/2022
							E2403	No Clinical Signs, Symptoms or Conditions	A070909	Oversensing	13/09/2022	13/09/2022
									A071302	Inappropriate/Inadequate Shock/Stimulation	13/09/2022	13/09/2022
									A090801	Signal Artifact/Noise	13/09/2022	13/09/2022
							F08	Hospitalization or Prolonged Hospitalization	A070909	Oversensing	13/09/2022	13/09/2022
									A071302	Inappropriate/Inadequate Shock/Stimulation	13/09/2022	13/09/2022
									A090801	Signal Artifact/Noise	13/09/2022	13/09/2022
							F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A070909	Oversensing	13/09/2022	13/09/2022
									A071302	Inappropriate/Inadequate Shock/Stimulation	13/09/2022	13/09/2022
									A090801	Signal Artifact/Noise	13/09/2022	13/09/2022
							F1905	Device Revision or Replacement	A070909	Oversensing	13/09/2022	13/09/2022
									A071302	Inappropriate/Inadequate Shock/Stimulation	13/09/2022	13/09/2022
									A090801	Signal Artifact/Noise	13/09/2022	13/09/2022
							F2203	Imaging Required	A070909	Oversensing	13/09/2022	13/09/2022
									A071302	Inappropriate/Inadequate Shock/Stimulation	13/09/2022	13/09/2022
									A090801	Signal Artifact/Noise	13/09/2022	13/09/2022
					CT2407	Active implantable	E2104	Electric Shock	A070909	Oversensing	13/09/2022	13/09/2022
									A071302	Inappropriate/Inadequate Shock/Stimulation	13/09/2022	13/09/2022
									A090801	Signal Artifact/Noise	13/09/2022	13/09/2022
							E2403	No Clinical Signs, Symptoms or Conditions	A070909	Oversensing	13/09/2022	13/09/2022
									A071302	Inappropriate/Inadequate Shock/Stimulation	13/09/2022	13/09/2022
									A090801	Signal Artifact/Noise	13/09/2022	13/09/2022
							F08	Hospitalization or Prolonged Hospitalization	A070909	Oversensing	13/09/2022	13/09/2022
									A071302	Inappropriate/Inadequate Shock/Stimulation	13/09/2022	13/09/2022
									A090801	Signal Artifact/Noise	13/09/2022	13/09/2022
							F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A070909	Oversensing	13/09/2022	13/09/2022
									A071302	Inappropriate/Inadequate Shock/Stimulation	13/09/2022	13/09/2022
									A090801	Signal Artifact/Noise	13/09/2022	13/09/2022
							F1905	Device Revision or Replacement	A070909	Oversensing	13/09/2022	13/09/2022
									A071302	Inappropriate/Inadequate Shock/Stimulation	13/09/2022	13/09/2022
									A090801	Signal Artifact/Noise	13/09/2022	13/09/2022
							F2203	Imaging Required	A070909	Oversensing	13/09/2022	13/09/2022

								A071302	Inappropriate/Inadequate Shock/Stimulation	13/09/2022	13/09/2022
								A090801	Signal Artifact/Noise	13/09/2022	13/09/2022
					CT257	Implantable cardiac pacemakers	E2104	Electric Shock	A070909	Oversensing	13/09/2022	13/09/2022
									A071302	Inappropriate/Inadequate Shock/Stimulation	13/09/2022	13/09/2022
									A090801	Signal Artifact/Noise	13/09/2022	13/09/2022
							E2403	No Clinical Signs, Symptoms or Conditions	A070909	Oversensing	13/09/2022	13/09/2022
									A071302	Inappropriate/Inadequate Shock/Stimulation	13/09/2022	13/09/2022
									A090801	Signal Artifact/Noise	13/09/2022	13/09/2022
							F08	Hospitalization or Prolonged Hospitalization	A070909	Oversensing	13/09/2022	13/09/2022
									A071302	Inappropriate/Inadequate Shock/Stimulation	13/09/2022	13/09/2022
									A090801	Signal Artifact/Noise	13/09/2022	13/09/2022
							F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A070909	Oversensing	13/09/2022	13/09/2022
									A071302	Inappropriate/Inadequate Shock/Stimulation	13/09/2022	13/09/2022
									A090801	Signal Artifact/Noise	13/09/2022	13/09/2022
							F1905	Device Revision or Replacement	A070909	Oversensing	13/09/2022	13/09/2022
									A071302	Inappropriate/Inadequate Shock/Stimulation	13/09/2022	13/09/2022
									A090801	Signal Artifact/Noise	13/09/2022	13/09/2022
							F2203	Imaging Required	A070909	Oversensing	13/09/2022	13/09/2022
									A071302	Inappropriate/Inadequate Shock/Stimulation	13/09/2022	13/09/2022
									A090801	Signal Artifact/Noise	13/09/2022	13/09/2022
					CT2727	Active implant devices/ancillaries	E2104	Electric Shock	A070909	Oversensing	13/09/2022	13/09/2022
									A071302	Inappropriate/Inadequate Shock/Stimulation	13/09/2022	13/09/2022
									A090801	Signal Artifact/Noise	13/09/2022	13/09/2022
							E2403	No Clinical Signs, Symptoms or Conditions	A070909	Oversensing	13/09/2022	13/09/2022
									A071302	Inappropriate/Inadequate Shock/Stimulation	13/09/2022	13/09/2022
									A090801	Signal Artifact/Noise	13/09/2022	13/09/2022
							F08	Hospitalization or Prolonged Hospitalization	A070909	Oversensing	13/09/2022	13/09/2022
									A071302	Inappropriate/Inadequate Shock/Stimulation	13/09/2022	13/09/2022
									A090801	Signal Artifact/Noise	13/09/2022	13/09/2022
							F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A070909	Oversensing	13/09/2022	13/09/2022
									A071302	Inappropriate/Inadequate Shock/Stimulation	13/09/2022	13/09/2022
									A090801	Signal Artifact/Noise	13/09/2022	13/09/2022
							F1905	Device Revision or Replacement	A070909	Oversensing	13/09/2022	13/09/2022
									A071302	Inappropriate/Inadequate Shock/Stimulation	13/09/2022	13/09/2022
									A090801	Signal Artifact/Noise	13/09/2022	13/09/2022
							F2203	Imaging Required	A070909	Oversensing	13/09/2022	13/09/2022
									A071302	Inappropriate/Inadequate Shock/Stimulation	13/09/2022	13/09/2022
									A090801	Signal Artifact/Noise	13/09/2022	13/09/2022
					CT2750	Implantable cardiac electrical stimulators and associated devices	E2104	Electric Shock	A070909	Oversensing	13/09/2022	13/09/2022
									A071302	Inappropriate/Inadequate Shock/Stimulation	13/09/2022	13/09/2022
									A090801	Signal Artifact/Noise	13/09/2022	13/09/2022
							E2403	No Clinical Signs, Symptoms or Conditions	A070909	Oversensing	13/09/2022	13/09/2022
									A071302	Inappropriate/Inadequate Shock/Stimulation	13/09/2022	13/09/2022
									A090801	Signal Artifact/Noise	13/09/2022	13/09/2022
							F08	Hospitalization or Prolonged Hospitalization	A070909	Oversensing	13/09/2022	13/09/2022
									A071302	Inappropriate/Inadequate Shock/Stimulation	13/09/2022	13/09/2022
									A090801	Signal Artifact/Noise	13/09/2022	13/09/2022
							F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A070909	Oversensing	13/09/2022	13/09/2022
									A071302	Inappropriate/Inadequate Shock/Stimulation	13/09/2022	13/09/2022
									A090801	Signal Artifact/Noise	13/09/2022	13/09/2022
							F1905	Device Revision or Replacement	A070909	Oversensing	13/09/2022	13/09/2022
									A071302	Inappropriate/Inadequate Shock/Stimulation	13/09/2022	13/09/2022
									A090801	Signal Artifact/Noise	13/09/2022	13/09/2022
							F2203	Imaging Required	A070909	Oversensing	13/09/2022	13/09/2022
									A071302	Inappropriate/Inadequate Shock/Stimulation	13/09/2022	13/09/2022
									A090801	Signal Artifact/Noise	13/09/2022	13/09/2022
					CT2866	Implantable cardiac electrical stimulators	E2104	Electric Shock	A070909	Oversensing	13/09/2022	13/09/2022
									A071302	Inappropriate/Inadequate Shock/Stimulation	13/09/2022	13/09/2022
									A090801	Signal Artifact/Noise	13/09/2022	13/09/2022
							E2403	No Clinical Signs, Symptoms or Conditions	A070909	Oversensing	13/09/2022	13/09/2022
									A071302	Inappropriate/Inadequate Shock/Stimulation	13/09/2022	13/09/2022
									A090801	Signal Artifact/Noise	13/09/2022	13/09/2022
							F08	Hospitalization or Prolonged Hospitalization	A070909	Oversensing	13/09/2022	13/09/2022
									A071302	Inappropriate/Inadequate Shock/Stimulation	13/09/2022	13/09/2022
									A090801	Signal Artifact/Noise	13/09/2022	13/09/2022
							F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A070909	Oversensing	13/09/2022	13/09/2022
									A071302	Inappropriate/Inadequate Shock/Stimulation	13/09/2022	13/09/2022
									A090801	Signal Artifact/Noise	13/09/2022	13/09/2022
							F1905	Device Revision or Replacement	A070909	Oversensing	13/09/2022	13/09/2022
									A071302	Inappropriate/Inadequate Shock/Stimulation	13/09/2022	13/09/2022
									A090801	Signal Artifact/Noise	13/09/2022	13/09/2022
							F2203	Imaging Required	A070909	Oversensing	13/09/2022	13/09/2022
									A071302	Inappropriate/Inadequate Shock/Stimulation	13/09/2022	13/09/2022
									A090801	Signal Artifact/Noise	13/09/2022	13/09/2022
					CT2985	Active	E2104	Electric Shock	A070909	Oversensing	13/09/2022	13/09/2022
									A071302	Inappropriate/Inadequate Shock/Stimulation	13/09/2022	13/09/2022
									A090801	Signal Artifact/Noise	13/09/2022	13/09/2022
							E2403	No Clinical Signs, Symptoms or Conditions	A070909	Oversensing	13/09/2022	13/09/2022
									A071302	Inappropriate/Inadequate Shock/Stimulation	13/09/2022	13/09/2022
									A090801	Signal Artifact/Noise	13/09/2022	13/09/2022
							F08	Hospitalization or Prolonged Hospitalization	A070909	Oversensing	13/09/2022	13/09/2022
									A071302	Inappropriate/Inadequate Shock/Stimulation	13/09/2022	13/09/2022
									A090801	Signal Artifact/Noise	13/09/2022	13/09/2022
							F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A070909	Oversensing	13/09/2022	13/09/2022
									A071302	Inappropriate/Inadequate Shock/Stimulation	13/09/2022	13/09/2022
									A090801	Signal Artifact/Noise	13/09/2022	13/09/2022
							F1905	Device Revision or Replacement	A070909	Oversensing	13/09/2022	13/09/2022
									A071302	Inappropriate/Inadequate Shock/Stimulation	13/09/2022	13/09/2022

									A090801	Signal Artifact/Noise	13/09/2022	13/09/2022
							F2203	Imaging Required	A070909	Oversensing	13/09/2022	13/09/2022
									A071302	Inappropriate/Inadequate Shock/Stimulation	13/09/2022	13/09/2022
									A090801	Signal Artifact/Noise	13/09/2022	13/09/2022
					CT314	Electrically powered	E2104	Electric Shock	A070909	Oversensing	13/09/2022	13/09/2022
									A071302	Inappropriate/Inadequate Shock/Stimulation	13/09/2022	13/09/2022
									A090801	Signal Artifact/Noise	13/09/2022	13/09/2022
							E2403	No Clinical Signs, Symptoms or Conditions	A070909	Oversensing	13/09/2022	13/09/2022
									A071302	Inappropriate/Inadequate Shock/Stimulation	13/09/2022	13/09/2022
									A090801	Signal Artifact/Noise	13/09/2022	13/09/2022
							F08	Hospitalization or Prolonged Hospitalization	A070909	Oversensing	13/09/2022	13/09/2022
									A071302	Inappropriate/Inadequate Shock/Stimulation	13/09/2022	13/09/2022
									A090801	Signal Artifact/Noise	13/09/2022	13/09/2022
							F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A070909	Oversensing	13/09/2022	13/09/2022
									A071302	Inappropriate/Inadequate Shock/Stimulation	13/09/2022	13/09/2022
									A090801	Signal Artifact/Noise	13/09/2022	13/09/2022
							F1905	Device Revision or Replacement	A070909	Oversensing	13/09/2022	13/09/2022
									A071302	Inappropriate/Inadequate Shock/Stimulation	13/09/2022	13/09/2022
									A090801	Signal Artifact/Noise	13/09/2022	13/09/2022
							F2203	Imaging Required	A070909	Oversensing	13/09/2022	13/09/2022
									A071302	Inappropriate/Inadequate Shock/Stimulation	13/09/2022	13/09/2022
									A090801	Signal Artifact/Noise	13/09/2022	13/09/2022
7807	2022	All other reportable incidents	33500	Pacing/defibrillation lead repair kit	CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A0722	Impedance Problem		
									A1301	Failure to Read Input Signal		
							F26	No Health Consequences or Impact	A0722	Impedance Problem		
									A1301	Failure to Read Input Signal		
7808	2022	All other reportable incidents	35852	Single-chamber implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E2104	Electric Shock	A070909	Oversensing		
									A071302	Inappropriate/Inadequate Shock/Stimulation		
							E2340	Wound Dehiscence	A070909	Oversensing		
									A071302	Inappropriate/Inadequate Shock/Stimulation		
							E2401	Insufficient Information	A070909	Oversensing		
									A071302	Inappropriate/Inadequate Shock/Stimulation		
							E2403	No Clinical Signs, Symptoms or Conditions	A070909	Oversensing		
									A071302	Inappropriate/Inadequate Shock/Stimulation		
							F08	Hospitalization or Prolonged Hospitalization	A070909	Oversensing		
									A071302	Inappropriate/Inadequate Shock/Stimulation		
							F1904	Device Repositioning	A070909	Oversensing		
									A071302	Inappropriate/Inadequate Shock/Stimulation		
7809	2022	All other reportable incidents	35223	Endocardial/interventricular septal pacing lead	CT1684	Cardiac pacemakers and associated devices	E0604	Cardiac Perforation	A24	No Apparent Device Problem		
							E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem		
							F1908	Prolonged Surgery	A24	No Apparent Device Problem		
7810	2022	All other reportable incidents	37265	Dual-chamber implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A070912	Incorrect Interpretation of Signal		
									A071302	Inappropriate/Inadequate Shock/Stimulation		
							F26	No Health Consequences or Impact	A070912	Incorrect Interpretation of Signal		
									A071302	Inappropriate/Inadequate Shock/Stimulation		
7811	2022	All other reportable incidents	35223	Endocardial/interventricular septal pacing lead	CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070901	Decreased Sensitivity		
									A070909	Oversensing		
									A070910	Undersensing		
							F26	No Health Consequences or Impact	A070901	Decreased Sensitivity		
									A070909	Oversensing		
									A070910	Undersensing		
7812	2022	All other reportable incidents	35223	Endocardial/interventricular septal pacing lead	CT1684	Cardiac pacemakers and associated devices	E2114	Perforation	A051201	Device Dislodged or Dislocated		
									A070101	Failure to Capture		
									A070102	High Capture Threshold		
									A0722	Impedance Problem		
							E2403	No Clinical Signs, Symptoms or Conditions	A051201	Device Dislodged or Dislocated		
									A070101	Failure to Capture		
									A070102	High Capture Threshold		
									A0722	Impedance Problem		
							F08	Hospitalization or Prolonged Hospitalization	A051201	Device Dislodged or Dislocated		
									A070101	Failure to Capture		
									A070102	High Capture Threshold		
									A0722	Impedance Problem		
							F1203	Life Threatening Illness or Injury	A051201	Device Dislodged or Dislocated		
									A070101	Failure to Capture		
									A070102	High Capture Threshold		
									A0722	Impedance Problem		
							F23	Unexpected Medical Intervention	A051201	Device Dislodged or Dislocated		
									A070101	Failure to Capture		
									A070102	High Capture Threshold		
									A0722	Impedance Problem		
7813	2022	All other reportable incidents	47265	Dual-chamber implantable pacemaker, rate-responsive	CT141	Electrodes/Leads and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	14/09/2022	14/09/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	14/09/2022	14/09/2022
					CT145	Cardiovascular devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	14/09/2022	14/09/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	14/09/2022	14/09/2022
					CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	14/09/2022	14/09/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	14/09/2022	14/09/2022
					CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	14/09/2022	14/09/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	14/09/2022	14/09/2022
					CT233	Surgical	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	14/09/2022	14/09/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	14/09/2022	14/09/2022
					CT2406	Implantable	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	14/09/2022	14/09/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	14/09/2022	14/09/2022
					CT2407	Active implantable	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	14/09/2022	14/09/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	14/09/2022	14/09/2022
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	14/09/2022	14/09/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	14/09/2022	14/09/2022
					CT2727	Active implant devices/ancillaries	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	14/09/2022	14/09/2022

							F1901	Additional Surgery	A24	No Apparent Device Problem	14/09/2022	14/09/2022
					CT2750	Implantable cardiac electrical stimulators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	14/09/2022	14/09/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	14/09/2022	14/09/2022
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	14/09/2022	14/09/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	14/09/2022	14/09/2022
					CT2985	Active	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	14/09/2022	14/09/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	14/09/2022	14/09/2022
					CT3001	Implantable metal	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	14/09/2022	14/09/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	14/09/2022	14/09/2022
					CT3002	Implantable synthetic polymer	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	14/09/2022	14/09/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	14/09/2022	14/09/2022
					CT3006	Implantable inorganic material	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	14/09/2022	14/09/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	14/09/2022	14/09/2022
					CT314	Electrically powered	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	14/09/2022	14/09/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	14/09/2022	14/09/2022
					CT321	Long-term surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	14/09/2022	14/09/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	14/09/2022	14/09/2022
					CT449	Generators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	14/09/2022	14/09/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	14/09/2022	14/09/2022
					CT981	Single use	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	14/09/2022	14/09/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	14/09/2022	14/09/2022
					CT983	Surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	14/09/2022	14/09/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	14/09/2022	14/09/2022
7814	2022		47265	Dual-chamber implantable pacemaker, rate-responsive	CT141	Electrodes/Leads and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	18/09/2022	18/09/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	18/09/2022	18/09/2022
					CT145	Cardiovascular devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	18/09/2022	18/09/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	18/09/2022	18/09/2022
					CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	18/09/2022	18/09/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	18/09/2022	18/09/2022
					CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	18/09/2022	18/09/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	18/09/2022	18/09/2022
					CT233	Surgical	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	18/09/2022	18/09/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	18/09/2022	18/09/2022
					CT2406	Implantable	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	18/09/2022	18/09/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	18/09/2022	18/09/2022
					CT2407	Active implantable	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	18/09/2022	18/09/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	18/09/2022	18/09/2022
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	18/09/2022	18/09/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	18/09/2022	18/09/2022
					CT2727	Active implant devices/ancillaries	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	18/09/2022	18/09/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	18/09/2022	18/09/2022
					CT2750	Implantable cardiac electrical stimulators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	18/09/2022	18/09/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	18/09/2022	18/09/2022
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	18/09/2022	18/09/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	18/09/2022	18/09/2022
					CT2985	Active	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	18/09/2022	18/09/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	18/09/2022	18/09/2022
					CT3001	Implantable metal	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	18/09/2022	18/09/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	18/09/2022	18/09/2022
					CT3002	Implantable synthetic polymer	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	18/09/2022	18/09/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	18/09/2022	18/09/2022
					CT3006	Implantable inorganic material	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	18/09/2022	18/09/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	18/09/2022	18/09/2022
					CT314	Electrically powered	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	18/09/2022	18/09/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	18/09/2022	18/09/2022
					CT321	Long-term surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	18/09/2022	18/09/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	18/09/2022	18/09/2022
					CT449	Generators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	18/09/2022	18/09/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	18/09/2022	18/09/2022
					CT981	Single use	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	18/09/2022	18/09/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	18/09/2022	18/09/2022
					CT983	Surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	18/09/2022	18/09/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	18/09/2022	18/09/2022
7815	2022	All other reportable incidents	35223	Endocardial/interventricular septal pacing lead	CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A040101	Fracture		
									A07	Electrical /Electronic Property Problem		
									A070910	Undersensing		
									A0722	Impedance Problem		
									A072201	High Impedance		
							F26	No Health Consequences or Impact	A040101	Fracture		
									A07	Electrical /Electronic Property Problem		
									A070910	Undersensing		
									A0722	Impedance Problem		
									A072201	High Impedance		
7816	2022	All other reportable incidents	47265	Dual-chamber implantable pacemaker, rate-responsive	CT141	Electrodes/Leads and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	18/09/2022	18/09/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	18/09/2022	18/09/2022
					CT145	Cardiovascular devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	18/09/2022	18/09/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	18/09/2022	18/09/2022
					CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	18/09/2022	18/09/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	18/09/2022	18/09/2022
					CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	18/09/2022	18/09/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	18/09/2022	18/09/2022
					CT233	Surgical	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	18/09/2022	18/09/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	18/09/2022	18/09/2022
					CT2406	Implantable	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	18/09/2022	18/09/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	18/09/2022	18/09/2022
					CT2407	Active implantable	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	18/09/2022	18/09/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	18/09/2022	18/09/2022
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	18/09/2022	18/09/2022

						F1901	Additional Surgery	A24	No Apparent Device Problem	18/09/2022	18/09/2022
					CT2727	Active implant devices/ancillaries	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	18/09/2022	18/09/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	18/09/2022	18/09/2022
					CT2750	Implantable cardiac electrical stimulators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	18/09/2022	18/09/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	18/09/2022	18/09/2022
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	18/09/2022	18/09/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	18/09/2022	18/09/2022
					CT2985	Active	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	18/09/2022	18/09/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	18/09/2022	18/09/2022
					CT3001	Implantable metal	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	18/09/2022	18/09/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	18/09/2022	18/09/2022
					CT3002	Implantable synthetic polymer	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	18/09/2022	18/09/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	18/09/2022	18/09/2022
					CT3006	Implantable inorganic material	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	18/09/2022	18/09/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	18/09/2022	18/09/2022
					CT314	Electrically powered	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	18/09/2022	18/09/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	18/09/2022	18/09/2022
					CT321	Long-term surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	18/09/2022	18/09/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	18/09/2022	18/09/2022
					CT449	Generators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	18/09/2022	18/09/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	18/09/2022	18/09/2022
					CT981	Single use	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	18/09/2022	18/09/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	18/09/2022	18/09/2022
					CT983	Surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	18/09/2022	18/09/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	18/09/2022	18/09/2022
7817	2022	All other reportable incidents	35223	Endocardial/interventricular septal pacing lead	CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A040101	Fracture		
									A0701	Capturing Problem		
									A070909	Oversensing		
									A23	Use of Device Problem		
							F23	Unexpected Medical Intervention	A040101	Fracture		
									A0701	Capturing Problem		
									A070909	Oversensing		
									A23	Use of Device Problem		
7818	2022	All other reportable incidents	47270	Cardiac resynchronization therapy implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A0712	Pacing Problem		
							F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A0712	Pacing Problem		
7819	2022	All other reportable incidents	35852	Single-chamber implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem	08/11/2022	08/11/2022
									A070504	Premature Discharge of Battery	08/11/2022	08/11/2022
									A0908	Incorrect, Inadequate or Imprecise Result or Readings	08/11/2022	08/11/2022
							F08	Hospitalization or Prolonged Hospitalization	A0705	Battery Problem	08/11/2022	08/11/2022
									A070504	Premature Discharge of Battery	08/11/2022	08/11/2022
									A0908	Incorrect, Inadequate or Imprecise Result or Readings	08/11/2022	08/11/2022
							F1905	Device Revision or Replacement	A0705	Battery Problem	08/11/2022	08/11/2022
									A070504	Premature Discharge of Battery	08/11/2022	08/11/2022
									A0908	Incorrect, Inadequate or Imprecise Result or Readings	08/11/2022	08/11/2022
							F26	No Health Consequences or Impact	A0705	Battery Problem	08/11/2022	08/11/2022
									A070504	Premature Discharge of Battery	08/11/2022	08/11/2022
									A0908	Incorrect, Inadequate or Imprecise Result or Readings	08/11/2022	08/11/2022
7820	2022	All other reportable incidents	37265	Dual-chamber implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E2104	Electric Shock	A0712	Pacing Problem		
									A071302	Inappropriate/Inadequate Shock/Stimulation		
							E2403	No Clinical Signs, Symptoms or Conditions	A0712	Pacing Problem		
									A071302	Inappropriate/Inadequate Shock/Stimulation		
							F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A0712	Pacing Problem		
									A071302	Inappropriate/Inadequate Shock/Stimulation		
7821	2022	All other reportable incidents	47270	Cardiac resynchronization therapy implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem		
									A072201	High Impedance		
							F26	No Health Consequences or Impact	A0705	Battery Problem		
									A072201	High Impedance		
7822	2022	All other reportable incidents	60789	Intracardiac pacemaker	CT130	Catheters and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem		
							F26	No Health Consequences or Impact	A0705	Battery Problem		
					CT145	Cardiovascular devices	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem		
							F26	No Health Consequences or Impact	A0705	Battery Problem		
					CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem		
							F26	No Health Consequences or Impact	A0705	Battery Problem		
					CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem		
							F26	No Health Consequences or Impact	A0705	Battery Problem		
					CT233	Surgical	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem		
							F26	No Health Consequences or Impact	A0705	Battery Problem		
					CT2406	Implantable	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem		
							F26	No Health Consequences or Impact	A0705	Battery Problem		
					CT2407	Active implantable	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem		
							F26	No Health Consequences or Impact	A0705	Battery Problem		
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem		
							F26	No Health Consequences or Impact	A0705	Battery Problem		
					CT2727	Active implant devices/ancillaries	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem		
							F26	No Health Consequences or Impact	A0705	Battery Problem		
					CT2750	Implantable cardiac electrical stimulators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem		
							F26	No Health Consequences or Impact	A0705	Battery Problem		
					CT2818	Catheterization/Cannulation/Fluid path devices	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem		
							F26	No Health Consequences or Impact	A0705	Battery Problem		
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem		
							F26	No Health Consequences or Impact	A0705	Battery Problem		
					CT2985	Active	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem		
							F26	No Health Consequences or Impact	A0705	Battery Problem		
					CT3001	Implantable metal	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem		
							F26	No Health Consequences or Impact	A0705	Battery Problem		

					CT3002	Implantable synthetic polymer	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem		
							F26	No Health Consequences or Impact	A0705	Battery Problem		
					CT3006	Implantable inorganic material	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem		
							F26	No Health Consequences or Impact	A0705	Battery Problem		
					CT314	Electrically powered	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem		
							F26	No Health Consequences or Impact	A0705	Battery Problem		
					CT321	Long-term surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem		
							F26	No Health Consequences or Impact	A0705	Battery Problem		
					CT981	Single use	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem		
							F26	No Health Consequences or Impact	A0705	Battery Problem		
					CT983	Surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem		
							F26	No Health Consequences or Impact	A0705	Battery Problem		
7823	2022	All other reportable incidents	47265	Dual-chamber implantable pacemaker, rate-responsive	CT141	Electrodes/Leads and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	24/02/2022	24/02/2022
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	24/02/2022	24/02/2022
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	24/02/2022	24/02/2022
					CT145	Cardiovascular devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	24/02/2022	24/02/2022
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	24/02/2022	24/02/2022
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	24/02/2022	24/02/2022
					CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	24/02/2022	24/02/2022
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	24/02/2022	24/02/2022
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	24/02/2022	24/02/2022
					CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	24/02/2022	24/02/2022
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	24/02/2022	24/02/2022
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	24/02/2022	24/02/2022
					CT233	Surgical	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	24/02/2022	24/02/2022
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	24/02/2022	24/02/2022
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	24/02/2022	24/02/2022
					CT2406	Implantable	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	24/02/2022	24/02/2022
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	24/02/2022	24/02/2022
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	24/02/2022	24/02/2022
					CT2407	Active implantable	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	24/02/2022	24/02/2022
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	24/02/2022	24/02/2022
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	24/02/2022	24/02/2022
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	24/02/2022	24/02/2022
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	24/02/2022	24/02/2022
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	24/02/2022	24/02/2022
					CT2727	Active implant devices/ancillaries	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	24/02/2022	24/02/2022
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	24/02/2022	24/02/2022
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	24/02/2022	24/02/2022
					CT2750	Implantable cardiac electrical stimulators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	24/02/2022	24/02/2022
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	24/02/2022	24/02/2022
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	24/02/2022	24/02/2022
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	24/02/2022	24/02/2022
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	24/02/2022	24/02/2022
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	24/02/2022	24/02/2022
					CT2985	Active	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	24/02/2022	24/02/2022
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	24/02/2022	24/02/2022
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	24/02/2022	24/02/2022
					CT3001	Implantable metal	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	24/02/2022	24/02/2022
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	24/02/2022	24/02/2022
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	24/02/2022	24/02/2022
					CT3002	Implantable synthetic polymer	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	24/02/2022	24/02/2022
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	24/02/2022	24/02/2022
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	24/02/2022	24/02/2022
					CT3006	Implantable inorganic material	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	24/02/2022	24/02/2022
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	24/02/2022	24/02/2022
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	24/02/2022	24/02/2022
					CT314	Electrically powered	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	24/02/2022	24/02/2022
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	24/02/2022	24/02/2022
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	24/02/2022	24/02/2022
					CT321	Long-term surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	24/02/2022	24/02/2022
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	24/02/2022	24/02/2022
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	24/02/2022	24/02/2022
					CT449	Generators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	24/02/2022	24/02/2022
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	24/02/2022	24/02/2022
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	24/02/2022	24/02/2022
					CT981	Single use	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	24/02/2022	24/02/2022
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	24/02/2022	24/02/2022
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	24/02/2022	24/02/2022
					CT983	Surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	24/02/2022	24/02/2022
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	24/02/2022	24/02/2022
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	24/02/2022	24/02/2022
7824	2022	All other reportable incidents	17882	Physiologic-monitoring defibrillation system	CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070803	Failure to Power Up		
							F27	Problem Identified During Non-Clinical Procedure	A070803	Failure to Power Up		
7825	2022	All other reportable incidents	17882	Physiologic-monitoring defibrillation system	CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070803	Failure to Power Up		
							F27	Problem Identified During Non-Clinical Procedure	A070803	Failure to Power Up		
7826	2022	All other reportable incidents	47265	Dual-chamber implantable pacemaker, rate-responsive	CT141	Electrodes/Leads and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A0203	Defective Device	14/03/2021	14/03/2021
							F24	Insufficient Health Impact Information	A0203	Defective Device	14/03/2021	14/03/2021
					CT145	Cardiovascular devices	E2403	No Clinical Signs, Symptoms or Conditions	A0203	Defective Device	14/03/2021	14/03/2021
							F24	Insufficient Health Impact Information	A0203	Defective Device	14/03/2021	14/03/2021
					CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A0203	Defective Device	14/03/2021	14/03/2021
							F24	Insufficient Health Impact Information	A0203	Defective Device	14/03/2021	14/03/2021
					CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A0203	Defective Device	14/03/2021	14/03/2021
							F24	Insufficient Health Impact Information	A0203	Defective Device	14/03/2021	14/03/2021
					CT233	Surgical	E2403	No Clinical Signs, Symptoms or Conditions	A0203	Defective Device	14/03/2021	14/03/2021
							F24	Insufficient Health Impact Information	A0203	Defective Device	14/03/2021	14/03/2021
					CT2406	Implantable	E2403	No Clinical Signs, Symptoms or Conditions	A0203	Defective Device	14/03/2021	14/03/2021
							F24	Insufficient Health Impact Information	A0203	Defective Device	14/03/2021	14/03/2021

					CT2407	Active implantable	E2403	No Clinical Signs, Symptoms or Conditions	A0203	Defective Device	14/03/2021	14/03/2021
							F24	Insufficient Health Impact Information	A0203	Defective Device	14/03/2021	14/03/2021
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A0203	Defective Device	14/03/2021	14/03/2021
							F24	Insufficient Health Impact Information	A0203	Defective Device	14/03/2021	14/03/2021
					CT2727	Active implant devices/ancillaries	E2403	No Clinical Signs, Symptoms or Conditions	A0203	Defective Device	14/03/2021	14/03/2021
							F24	Insufficient Health Impact Information	A0203	Defective Device	14/03/2021	14/03/2021
					CT2750	Implantable cardiac electrical stimulators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A0203	Defective Device	14/03/2021	14/03/2021
							F24	Insufficient Health Impact Information	A0203	Defective Device	14/03/2021	14/03/2021
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A0203	Defective Device	14/03/2021	14/03/2021
							F24	Insufficient Health Impact Information	A0203	Defective Device	14/03/2021	14/03/2021
					CT2985	Active	E2403	No Clinical Signs, Symptoms or Conditions	A0203	Defective Device	14/03/2021	14/03/2021
							F24	Insufficient Health Impact Information	A0203	Defective Device	14/03/2021	14/03/2021
					CT3001	Implantable metal	E2403	No Clinical Signs, Symptoms or Conditions	A0203	Defective Device	14/03/2021	14/03/2021
							F24	Insufficient Health Impact Information	A0203	Defective Device	14/03/2021	14/03/2021
					CT3002	Implantable synthetic polymer	E2403	No Clinical Signs, Symptoms or Conditions	A0203	Defective Device	14/03/2021	14/03/2021
							F24	Insufficient Health Impact Information	A0203	Defective Device	14/03/2021	14/03/2021
					CT3006	Implantable inorganic material	E2403	No Clinical Signs, Symptoms or Conditions	A0203	Defective Device	14/03/2021	14/03/2021
							F24	Insufficient Health Impact Information	A0203	Defective Device	14/03/2021	14/03/2021
					CT314	Electrically powered	E2403	No Clinical Signs, Symptoms or Conditions	A0203	Defective Device	14/03/2021	14/03/2021
							F24	Insufficient Health Impact Information	A0203	Defective Device	14/03/2021	14/03/2021
					CT321	Long-term surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A0203	Defective Device	14/03/2021	14/03/2021
							F24	Insufficient Health Impact Information	A0203	Defective Device	14/03/2021	14/03/2021
					CT449	Generators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A0203	Defective Device	14/03/2021	14/03/2021
							F24	Insufficient Health Impact Information	A0203	Defective Device	14/03/2021	14/03/2021
					CT981	Single use	E2403	No Clinical Signs, Symptoms or Conditions	A0203	Defective Device	14/03/2021	14/03/2021
							F24	Insufficient Health Impact Information	A0203	Defective Device	14/03/2021	14/03/2021
					CT983	Surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A0203	Defective Device	14/03/2021	14/03/2021
							F24	Insufficient Health Impact Information	A0203	Defective Device	14/03/2021	14/03/2021
7827	2022	All other reportable incidents	35852	Single-chamber implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E2104	Electric Shock	A070909	Oversensing		
									A070910	Undersensing		
									A071302	Inappropriate/Inadequate Shock/Stimulation		
									A090801	Signal Artifact/Noise		
									A1301	Failure to Read Input Signal		
							E2403	No Clinical Signs, Symptoms or Conditions	A070909	Oversensing		
									A070910	Undersensing		
									A071302	Inappropriate/Inadequate Shock/Stimulation		
									A090801	Signal Artifact/Noise		
									A1301	Failure to Read Input Signal		
							F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A070909	Oversensing		
									A070910	Undersensing		
									A071302	Inappropriate/Inadequate Shock/Stimulation		
									A090801	Signal Artifact/Noise		
									A1301	Failure to Read Input Signal		
7828	2022	All other reportable incidents	47265	Dual-chamber implantable pacemaker, rate-responsive	CT141	Electrodes/Leads and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	10/12/2021	10/12/2021
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	10/12/2021	10/12/2021
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	10/12/2021	10/12/2021
					CT145	Cardiovascular devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	10/12/2021	10/12/2021
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	10/12/2021	10/12/2021
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	10/12/2021	10/12/2021
					CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	10/12/2021	10/12/2021
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	10/12/2021	10/12/2021
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	10/12/2021	10/12/2021
					CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	10/12/2021	10/12/2021
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	10/12/2021	10/12/2021
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	10/12/2021	10/12/2021
					CT233	Surgical	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	10/12/2021	10/12/2021
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	10/12/2021	10/12/2021
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	10/12/2021	10/12/2021
					CT2406	Implantable	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	10/12/2021	10/12/2021
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	10/12/2021	10/12/2021
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	10/12/2021	10/12/2021
					CT2407	Active implantable	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	10/12/2021	10/12/2021
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	10/12/2021	10/12/2021
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	10/12/2021	10/12/2021
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	10/12/2021	10/12/2021
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	10/12/2021	10/12/2021
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	10/12/2021	10/12/2021
					CT2727	Active implant devices/ancillaries	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	10/12/2021	10/12/2021
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	10/12/2021	10/12/2021
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	10/12/2021	10/12/2021
					CT2750	Implantable cardiac electrical stimulators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	10/12/2021	10/12/2021
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	10/12/2021	10/12/2021
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	10/12/2021	10/12/2021
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	10/12/2021	10/12/2021
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	10/12/2021	10/12/2021
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	10/12/2021	10/12/2021
					CT2985	Active	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	10/12/2021	10/12/2021
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	10/12/2021	10/12/2021
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	10/12/2021	10/12/2021
					CT3001	Implantable metal	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	10/12/2021	10/12/2021
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	10/12/2021	10/12/2021
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	10/12/2021	10/12/2021
					CT3002	Implantable synthetic polymer	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	10/12/2021	10/12/2021
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	10/12/2021	10/12/2021
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	10/12/2021	10/12/2021
					CT3006	Implantable inorganic material	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	10/12/2021	10/12/2021
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	10/12/2021	10/12/2021
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	10/12/2021	10/12/2021

					CT314	Electrically powered	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	10/12/2021	10/12/2021
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	10/12/2021	10/12/2021
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	10/12/2021	10/12/2021
					CT321	Long-term surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	10/12/2021	10/12/2021
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	10/12/2021	10/12/2021
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	10/12/2021	10/12/2021
					CT449	Generators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	10/12/2021	10/12/2021
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	10/12/2021	10/12/2021
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	10/12/2021	10/12/2021
					CT981	Single use	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	10/12/2021	10/12/2021
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	10/12/2021	10/12/2021
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	10/12/2021	10/12/2021
					CT983	Surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	10/12/2021	10/12/2021
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	10/12/2021	10/12/2021
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	10/12/2021	10/12/2021
7829	2022	All other reportable incidents	35852	Single-chamber implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E2104	Electric Shock	A0508	Noise, Audible		
									A071302	Inappropriate/Inadequate Shock/Stimulation		
									A072201	High Impedance		
									A0908	Incorrect, Inadequate or Imprecise Result or Readings		
							E2403	No Clinical Signs, Symptoms or Conditions	A0508	Noise, Audible		
									A071302	Inappropriate/Inadequate Shock/Stimulation		
									A072201	High Impedance		
									A0908	Incorrect, Inadequate or Imprecise Result or Readings		
							F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A0508	Noise, Audible		
									A071302	Inappropriate/Inadequate Shock/Stimulation		
									A072201	High Impedance		
									A0908	Incorrect, Inadequate or Imprecise Result or Readings		
7830	2022	All other reportable incidents	35852	Single-chamber implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	23/09/2022	23/09/2022
							F1901	Additional Surgery	A070504	Premature Discharge of Battery	23/09/2022	23/09/2022
7831	2022	All other reportable incidents	17882	Physiologic-monitoring defibrillation system	CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A07	Electrical /Electronic Property Problem		
							F27	Problem Identified During Non-Clinical Procedure	A07	Electrical /Electronic Property Problem		
7832	2022	All other reportable incidents	35223	Endocardial/interventricular septal pacing lead	CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070102	High Capture Threshold	23/09/2022	23/09/2022
									A070103	Intermittent Capture	23/09/2022	23/09/2022
									A072201	High Impedance	23/09/2022	23/09/2022
									A15	Activation, Positioning or Separation Problem	23/09/2022	23/09/2022
							F08	Hospitalization or Prolonged Hospitalization	A070102	High Capture Threshold	23/09/2022	23/09/2022
									A070103	Intermittent Capture	23/09/2022	23/09/2022
									A072201	High Impedance	23/09/2022	23/09/2022
									A15	Activation, Positioning or Separation Problem	23/09/2022	23/09/2022
							F12	Serious Injury/ Illness/ Impairment	A070102	High Capture Threshold	23/09/2022	23/09/2022
									A070103	Intermittent Capture	23/09/2022	23/09/2022
									A072201	High Impedance	23/09/2022	23/09/2022
									A15	Activation, Positioning or Separation Problem	23/09/2022	23/09/2022
							F1905	Device Revision or Replacement	A070102	High Capture Threshold	23/09/2022	23/09/2022
									A070103	Intermittent Capture	23/09/2022	23/09/2022
									A072201	High Impedance	23/09/2022	23/09/2022
									A15	Activation, Positioning or Separation Problem	23/09/2022	23/09/2022
7833	2022	All other reportable incidents	35852	Single-chamber implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E2104	Electric Shock	A070909	Oversensing		
									A071302	Inappropriate/Inadequate Shock/Stimulation		
							E2403	No Clinical Signs, Symptoms or Conditions	A070909	Oversensing		
									A071302	Inappropriate/Inadequate Shock/Stimulation		
							F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A070909	Oversensing		
									A071302	Inappropriate/Inadequate Shock/Stimulation		
7834	2022	All other reportable incidents	35852	Single-chamber implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem	12/10/2022	12/10/2022
									A070504	Premature Discharge of Battery	12/10/2022	12/10/2022
							F08	Hospitalization or Prolonged Hospitalization	A0705	Battery Problem	12/10/2022	12/10/2022
									A070504	Premature Discharge of Battery	12/10/2022	12/10/2022
							F1905	Device Revision or Replacement	A0705	Battery Problem	12/10/2022	12/10/2022
									A070504	Premature Discharge of Battery	12/10/2022	12/10/2022
7835	2022	All other reportable incidents	47265	Dual-chamber implantable pacemaker, rate-responsive	CT141	Electrodes/Leads and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A090804	False Positive Result		
							F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A090804	False Positive Result		
					CT145	Cardiovascular devices	E2403	No Clinical Signs, Symptoms or Conditions	A090804	False Positive Result		
							F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A090804	False Positive Result		
					CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A090804	False Positive Result		
							F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A090804	False Positive Result		
					CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A090804	False Positive Result		
							F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A090804	False Positive Result		
					CT233	Surgical	E2403	No Clinical Signs, Symptoms or Conditions	A090804	False Positive Result		
							F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A090804	False Positive Result		
					CT2406	Implantable	E2403	No Clinical Signs, Symptoms or Conditions	A090804	False Positive Result		
							F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A090804	False Positive Result		
					CT2407	Active implantable	E2403	No Clinical Signs, Symptoms or Conditions	A090804	False Positive Result		
							F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A090804	False Positive Result		
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A090804	False Positive Result		
							F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A090804	False Positive Result		
					CT2727	Active implant devices/ancillaries	E2403	No Clinical Signs, Symptoms or Conditions	A090804	False Positive Result		
							F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A090804	False Positive Result		
					CT2750	Implantable cardiac electrical stimulators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A090804	False Positive Result		
							F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A090804	False Positive Result		
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A090804	False Positive Result		
							F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A090804	False Positive Result		
					CT2985	Active	E2403	No Clinical Signs, Symptoms or Conditions	A090804	False Positive Result		
							F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A090804	False Positive Result		
					CT3001	Implantable metal	E2403	No Clinical Signs, Symptoms or Conditions	A090804	False Positive Result		
							F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A090804	False Positive Result		

					CT3002	Implantable synthetic polymer	E2403	No Clinical Signs, Symptoms or Conditions	A090804	False Positive Result		
							F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A090804	False Positive Result		
					CT3006	Implantable inorganic material	E2403	No Clinical Signs, Symptoms or Conditions	A090804	False Positive Result		
							F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A090804	False Positive Result		
					CT314	Electrically powered	E2403	No Clinical Signs, Symptoms or Conditions	A090804	False Positive Result		
							F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A090804	False Positive Result		
					CT321	Long-term surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A090804	False Positive Result		
							F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A090804	False Positive Result		
					CT449	Generators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A090804	False Positive Result		
							F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A090804	False Positive Result		
					CT981	Single use	E2403	No Clinical Signs, Symptoms or Conditions	A090804	False Positive Result		
							F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A090804	False Positive Result		
					CT983	Surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A090804	False Positive Result		
							F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A090804	False Positive Result		
7836	2022	All other reportable incidents	37265	Dual-chamber implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A072201	High Impedance		
							F26	No Health Consequences or Impact	A072201	High Impedance		
7837	2022	All other reportable incidents	47265	Dual-chamber implantable pacemaker, rate-responsive	CT141	Electrodes/Leads and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	15/09/2022	15/09/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	15/09/2022	15/09/2022
					CT145	Cardiovascular devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	15/09/2022	15/09/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	15/09/2022	15/09/2022
					CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	15/09/2022	15/09/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	15/09/2022	15/09/2022
					CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	15/09/2022	15/09/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	15/09/2022	15/09/2022
					CT233	Surgical	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	15/09/2022	15/09/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	15/09/2022	15/09/2022
					CT2406	Implantable	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	15/09/2022	15/09/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	15/09/2022	15/09/2022
					CT2407	Active implantable	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	15/09/2022	15/09/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	15/09/2022	15/09/2022
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	15/09/2022	15/09/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	15/09/2022	15/09/2022
					CT2727	Active implant devices/ancillaries	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	15/09/2022	15/09/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	15/09/2022	15/09/2022
					CT2750	Implantable cardiac electrical stimulators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	15/09/2022	15/09/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	15/09/2022	15/09/2022
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	15/09/2022	15/09/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	15/09/2022	15/09/2022
					CT2985	Active	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	15/09/2022	15/09/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	15/09/2022	15/09/2022
					CT3001	Implantable metal	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	15/09/2022	15/09/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	15/09/2022	15/09/2022
					CT3002	Implantable synthetic polymer	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	15/09/2022	15/09/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	15/09/2022	15/09/2022
					CT3006	Implantable inorganic material	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	15/09/2022	15/09/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	15/09/2022	15/09/2022
					CT314	Electrically powered	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	15/09/2022	15/09/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	15/09/2022	15/09/2022
					CT321	Long-term surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	15/09/2022	15/09/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	15/09/2022	15/09/2022
					CT449	Generators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	15/09/2022	15/09/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	15/09/2022	15/09/2022
					CT981	Single use	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	15/09/2022	15/09/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	15/09/2022	15/09/2022
					CT983	Surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	15/09/2022	15/09/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	15/09/2022	15/09/2022
7838	2022	All other reportable incidents	47265	Dual-chamber implantable pacemaker, rate-responsive	CT141	Electrodes/Leads and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	21/09/2022	21/09/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	21/09/2022	21/09/2022
					CT145	Cardiovascular devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	21/09/2022	21/09/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	21/09/2022	21/09/2022
					CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	21/09/2022	21/09/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	21/09/2022	21/09/2022
					CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	21/09/2022	21/09/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	21/09/2022	21/09/2022
					CT233	Surgical	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	21/09/2022	21/09/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	21/09/2022	21/09/2022
					CT2406	Implantable	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	21/09/2022	21/09/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	21/09/2022	21/09/2022
					CT2407	Active implantable	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	21/09/2022	21/09/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	21/09/2022	21/09/2022
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	21/09/2022	21/09/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	21/09/2022	21/09/2022
					CT2727	Active implant devices/ancillaries	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	21/09/2022	21/09/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	21/09/2022	21/09/2022
					CT2750	Implantable cardiac electrical stimulators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	21/09/2022	21/09/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	21/09/2022	21/09/2022
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	21/09/2022	21/09/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	21/09/2022	21/09/2022
					CT2985	Active	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	21/09/2022	21/09/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	21/09/2022	21/09/2022
					CT3001	Implantable metal	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	21/09/2022	21/09/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	21/09/2022	21/09/2022
					CT3002	Implantable synthetic polymer	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	21/09/2022	21/09/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	21/09/2022	21/09/2022
					CT3006	Implantable inorganic material	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	21/09/2022	21/09/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	21/09/2022	21/09/2022
					CT314	Electrically powered	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	21/09/2022	21/09/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	21/09/2022	21/09/2022

					CT321	Long-term surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	21/09/2022	21/09/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	21/09/2022	21/09/2022
					CT449	Generators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	21/09/2022	21/09/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	21/09/2022	21/09/2022
					CT981	Single use	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	21/09/2022	21/09/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	21/09/2022	21/09/2022
					CT983	Surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	21/09/2022	21/09/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	21/09/2022	21/09/2022
7839	2022		35223	Endocardial/interventricular septal pacing lead	CT1684	Cardiac pacemakers and associated devices	E0601	Arrhythmia	A150204	Failure to Advance		
									A23	Use of Device Problem		
							E2403	No Clinical Signs, Symptoms or Conditions	A150204	Failure to Advance		
									A23	Use of Device Problem		
							F1908	Prolonged Surgery	A150204	Failure to Advance		
									A23	Use of Device Problem		
7840	2022	All other reportable incidents	47265	Dual-chamber implantable pacemaker, rate-responsive	CT141	Electrodes/Leads and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	18/09/2022	18/09/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	18/09/2022	18/09/2022
					CT145	Cardiovascular devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	18/09/2022	18/09/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	18/09/2022	18/09/2022
					CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	18/09/2022	18/09/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	18/09/2022	18/09/2022
					CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	18/09/2022	18/09/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	18/09/2022	18/09/2022
					CT233	Surgical	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	18/09/2022	18/09/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	18/09/2022	18/09/2022
					CT2406	Implantable	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	18/09/2022	18/09/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	18/09/2022	18/09/2022
					CT2407	Active implantable	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	18/09/2022	18/09/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	18/09/2022	18/09/2022
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	18/09/2022	18/09/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	18/09/2022	18/09/2022
					CT2727	Active implant devices/ancillaries	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	18/09/2022	18/09/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	18/09/2022	18/09/2022
					CT2750	Implantable cardiac electrical stimulators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	18/09/2022	18/09/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	18/09/2022	18/09/2022
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	18/09/2022	18/09/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	18/09/2022	18/09/2022
					CT2985	Active	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	18/09/2022	18/09/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	18/09/2022	18/09/2022
					CT3001	Implantable metal	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	18/09/2022	18/09/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	18/09/2022	18/09/2022
					CT3002	Implantable synthetic polymer	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	18/09/2022	18/09/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	18/09/2022	18/09/2022
					CT3006	Implantable inorganic material	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	18/09/2022	18/09/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	18/09/2022	18/09/2022
					CT314	Electrically powered	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	18/09/2022	18/09/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	18/09/2022	18/09/2022
					CT321	Long-term surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	18/09/2022	18/09/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	18/09/2022	18/09/2022
					CT449	Generators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	18/09/2022	18/09/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	18/09/2022	18/09/2022
					CT981	Single use	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	18/09/2022	18/09/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	18/09/2022	18/09/2022
					CT983	Surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	18/09/2022	18/09/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	18/09/2022	18/09/2022
7841	2022	All other reportable incidents	37265	Dual-chamber implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	26/09/2022	26/09/2022
							F1901	Additional Surgery	A070504	Premature Discharge of Battery	26/09/2022	26/09/2022
7842	2022	All other reportable incidents	17882	Physiologic-monitoring defibrillation system	CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070803	Failure to Power Up		
							F27	Problem Identified During Non-Clinical Procedure	A070803	Failure to Power Up		
7843	2022	All other reportable incidents	47265	Dual-chamber implantable pacemaker, rate-responsive	CT141	Electrodes/Leads and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT145	Cardiovascular devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT233	Surgical	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT2406	Implantable	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT2407	Active implantable	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT2727	Active implant devices/ancillaries	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT2750	Implantable cardiac electrical stimulators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT2985	Active	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT3001	Implantable metal	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT3002	Implantable synthetic polymer	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT3006	Implantable inorganic material	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		

					CT314	Electrically powered	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT321	Long-term surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT449	Generators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT981	Single use	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT983	Surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
7844	2022	All other reportable incidents	17882	Physiologic-monitoring defibrillation system	CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070803	Failure to Power Up		
									A071302	Inappropriate/Inadequate Shock/Stimulation		
							F27	Problem Identified During Non-Clinical Procedure	A070803	Failure to Power Up		
									A071302	Inappropriate/Inadequate Shock/Stimulation		
7845	2022	All other reportable incidents	17882	Physiologic-monitoring defibrillation system	CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070803	Failure to Power Up		
									A071302	Inappropriate/Inadequate Shock/Stimulation		
							F27	Problem Identified During Non-Clinical Procedure	A070803	Failure to Power Up		
									A071302	Inappropriate/Inadequate Shock/Stimulation		
7846	2022	All other reportable incidents	36241	Pacemaker/defibrillator lead	CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A072201	High Impedance		
							F26	No Health Consequences or Impact	A072201	High Impedance		
7847	2022	All other reportable incidents	47263	Cardiac resynchronization therapy implantable pacemaker	CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A0908	Incorrect, Inadequate or Imprecise Result or Readings	30/11/2022	30/11/2022
									A160503	Inappropriate or Unexpected Reset	30/11/2022	30/11/2022
							F08	Hospitalization or Prolonged Hospitalization	A0908	Incorrect, Inadequate or Imprecise Result or Readings	30/11/2022	30/11/2022
									A160503	Inappropriate or Unexpected Reset	30/11/2022	30/11/2022
							F1905	Device Revision or Replacement	A0908	Incorrect, Inadequate or Imprecise Result or Readings	30/11/2022	30/11/2022
									A160503	Inappropriate or Unexpected Reset	30/11/2022	30/11/2022
					CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A0908	Incorrect, Inadequate or Imprecise Result or Readings	30/11/2022	30/11/2022
									A160503	Inappropriate or Unexpected Reset	30/11/2022	30/11/2022
							F08	Hospitalization or Prolonged Hospitalization	A0908	Incorrect, Inadequate or Imprecise Result or Readings	30/11/2022	30/11/2022
									A160503	Inappropriate or Unexpected Reset	30/11/2022	30/11/2022
							F1905	Device Revision or Replacement	A0908	Incorrect, Inadequate or Imprecise Result or Readings	30/11/2022	30/11/2022
									A160503	Inappropriate or Unexpected Reset	30/11/2022	30/11/2022
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A0908	Incorrect, Inadequate or Imprecise Result or Readings	30/11/2022	30/11/2022
									A160503	Inappropriate or Unexpected Reset	30/11/2022	30/11/2022
							F08	Hospitalization or Prolonged Hospitalization	A0908	Incorrect, Inadequate or Imprecise Result or Readings	30/11/2022	30/11/2022
									A160503	Inappropriate or Unexpected Reset	30/11/2022	30/11/2022
							F1905	Device Revision or Replacement	A0908	Incorrect, Inadequate or Imprecise Result or Readings	30/11/2022	30/11/2022
									A160503	Inappropriate or Unexpected Reset	30/11/2022	30/11/2022
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A0908	Incorrect, Inadequate or Imprecise Result or Readings	30/11/2022	30/11/2022
									A160503	Inappropriate or Unexpected Reset	30/11/2022	30/11/2022
							F08	Hospitalization or Prolonged Hospitalization	A0908	Incorrect, Inadequate or Imprecise Result or Readings	30/11/2022	30/11/2022
									A160503	Inappropriate or Unexpected Reset	30/11/2022	30/11/2022
							F1905	Device Revision or Replacement	A0908	Incorrect, Inadequate or Imprecise Result or Readings	30/11/2022	30/11/2022
									A160503	Inappropriate or Unexpected Reset	30/11/2022	30/11/2022
7848	2022	All other reportable incidents	47265	Dual-chamber implantable pacemaker, rate-responsive	CT141	Electrodes/Leads and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	09/09/2022	09/09/2022
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	09/09/2022	09/09/2022
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	09/09/2022	09/09/2022
					CT145	Cardiovascular devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	09/09/2022	09/09/2022
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	09/09/2022	09/09/2022
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	09/09/2022	09/09/2022
					CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	09/09/2022	09/09/2022
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	09/09/2022	09/09/2022
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	09/09/2022	09/09/2022
					CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	09/09/2022	09/09/2022
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	09/09/2022	09/09/2022
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	09/09/2022	09/09/2022
					CT233	Surgical	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	09/09/2022	09/09/2022
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	09/09/2022	09/09/2022
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	09/09/2022	09/09/2022
					CT2406	Implantable	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	09/09/2022	09/09/2022
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	09/09/2022	09/09/2022
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	09/09/2022	09/09/2022
					CT2407	Active implantable	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	09/09/2022	09/09/2022
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	09/09/2022	09/09/2022
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	09/09/2022	09/09/2022
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	09/09/2022	09/09/2022
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	09/09/2022	09/09/2022
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	09/09/2022	09/09/2022
					CT2727	Active implant devices/ancillaries	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	09/09/2022	09/09/2022
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	09/09/2022	09/09/2022
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	09/09/2022	09/09/2022
					CT2750	Implantable cardiac electrical stimulators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	09/09/2022	09/09/2022
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	09/09/2022	09/09/2022
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	09/09/2022	09/09/2022
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	09/09/2022	09/09/2022
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	09/09/2022	09/09/2022
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	09/09/2022	09/09/2022
					CT2985	Active	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	09/09/2022	09/09/2022
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	09/09/2022	09/09/2022

						F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	09/09/2022	09/09/2022
					CT3001	Implantable metal	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	09/09/2022	09/09/2022
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	09/09/2022	09/09/2022
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	09/09/2022	09/09/2022
					CT3002	Implantable synthetic polymer	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	09/09/2022	09/09/2022
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	09/09/2022	09/09/2022
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	09/09/2022	09/09/2022
					CT3006	Implantable inorganic material	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	09/09/2022	09/09/2022
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	09/09/2022	09/09/2022
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	09/09/2022	09/09/2022
					CT314	Electrically powered	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	09/09/2022	09/09/2022
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	09/09/2022	09/09/2022
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	09/09/2022	09/09/2022
					CT321	Long-term surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	09/09/2022	09/09/2022
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	09/09/2022	09/09/2022
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	09/09/2022	09/09/2022
					CT449	Generators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	09/09/2022	09/09/2022
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	09/09/2022	09/09/2022
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	09/09/2022	09/09/2022
					CT981	Single use	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	09/09/2022	09/09/2022
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	09/09/2022	09/09/2022
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	09/09/2022	09/09/2022
					CT983	Surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	09/09/2022	09/09/2022
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	09/09/2022	09/09/2022
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	09/09/2022	09/09/2022
7849	2022	All other reportable incidents	35852	Single-chamber implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	23/02/2022	23/02/2022
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	23/02/2022	23/02/2022
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	23/02/2022	23/02/2022
7850	2022	All other reportable incidents	37265	Dual-chamber implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A072201	High Impedance		
							F26	No Health Consequences or Impact	A072201	High Impedance		
7851	2022	All other reportable incidents	37265	Dual-chamber implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A070909	Oversensing		
									A090801	Signal Artifact/Noise		
							F26	No Health Consequences or Impact	A070909	Oversensing		
									A090801	Signal Artifact/Noise		
7852	2022	All other reportable incidents	47265	Dual-chamber implantable pacemaker, rate-responsive	CT141	Electrodes/Leads and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	24/09/2022	24/09/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	24/09/2022	24/09/2022
					CT145	Cardiovascular devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	24/09/2022	24/09/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	24/09/2022	24/09/2022
					CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	24/09/2022	24/09/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	24/09/2022	24/09/2022
					CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	24/09/2022	24/09/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	24/09/2022	24/09/2022
					CT233	Surgical	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	24/09/2022	24/09/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	24/09/2022	24/09/2022
					CT2406	Implantable	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	24/09/2022	24/09/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	24/09/2022	24/09/2022
					CT2407	Active implantable	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	24/09/2022	24/09/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	24/09/2022	24/09/2022
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	24/09/2022	24/09/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	24/09/2022	24/09/2022
					CT2727	Active implant devices/ancillaries	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	24/09/2022	24/09/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	24/09/2022	24/09/2022
					CT2750	Implantable cardiac electrical stimulators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	24/09/2022	24/09/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	24/09/2022	24/09/2022
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	24/09/2022	24/09/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	24/09/2022	24/09/2022
					CT2985	Active	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	24/09/2022	24/09/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	24/09/2022	24/09/2022
					CT3001	Implantable metal	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	24/09/2022	24/09/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	24/09/2022	24/09/2022
					CT3002	Implantable synthetic polymer	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	24/09/2022	24/09/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	24/09/2022	24/09/2022
					CT3006	Implantable inorganic material	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	24/09/2022	24/09/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	24/09/2022	24/09/2022
					CT314	Electrically powered	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	24/09/2022	24/09/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	24/09/2022	24/09/2022
					CT321	Long-term surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	24/09/2022	24/09/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	24/09/2022	24/09/2022
					CT449	Generators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	24/09/2022	24/09/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	24/09/2022	24/09/2022
					CT981	Single use	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	24/09/2022	24/09/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	24/09/2022	24/09/2022
					CT983	Surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	24/09/2022	24/09/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	24/09/2022	24/09/2022
7853	2022		47265	Dual-chamber implantable pacemaker, rate-responsive	CT141	Electrodes/Leads and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	21/09/2022	21/09/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	21/09/2022	21/09/2022
					CT145	Cardiovascular devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	21/09/2022	21/09/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	21/09/2022	21/09/2022
					CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	21/09/2022	21/09/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	21/09/2022	21/09/2022
					CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	21/09/2022	21/09/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	21/09/2022	21/09/2022
					CT233	Surgical	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	21/09/2022	21/09/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	21/09/2022	21/09/2022
					CT2406	Implantable	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	21/09/2022	21/09/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	21/09/2022	21/09/2022
					CT2407	Active implantable	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	21/09/2022	21/09/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	21/09/2022	21/09/2022

					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	21/09/2022	21/09/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	21/09/2022	21/09/2022
					CT2727	Active implant devices/ancillaries	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	21/09/2022	21/09/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	21/09/2022	21/09/2022
					CT2750	Implantable cardiac electrical stimulators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	21/09/2022	21/09/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	21/09/2022	21/09/2022
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	21/09/2022	21/09/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	21/09/2022	21/09/2022
					CT2985	Active	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	21/09/2022	21/09/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	21/09/2022	21/09/2022
					CT3001	Implantable metal	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	21/09/2022	21/09/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	21/09/2022	21/09/2022
					CT3002	Implantable synthetic polymer	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	21/09/2022	21/09/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	21/09/2022	21/09/2022
					CT3006	Implantable inorganic material	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	21/09/2022	21/09/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	21/09/2022	21/09/2022
					CT314	Electrically powered	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	21/09/2022	21/09/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	21/09/2022	21/09/2022
					CT321	Long-term surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	21/09/2022	21/09/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	21/09/2022	21/09/2022
					CT449	Generators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	21/09/2022	21/09/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	21/09/2022	21/09/2022
					CT981	Single use	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	21/09/2022	21/09/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	21/09/2022	21/09/2022
					CT983	Surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	21/09/2022	21/09/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	21/09/2022	21/09/2022
7854	2022	All other reportable incidents	47265	Dual-chamber implantable pacemaker, rate-responsive	CT141	Electrodes/Leads and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A071102	Failure to Interrogate	23/09/2022	23/09/2022
							F1901	Additional Surgery	A071102	Failure to Interrogate	23/09/2022	23/09/2022
					CT145	Cardiovascular devices	E2403	No Clinical Signs, Symptoms or Conditions	A071102	Failure to Interrogate	23/09/2022	23/09/2022
							F1901	Additional Surgery	A071102	Failure to Interrogate	23/09/2022	23/09/2022
					CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A071102	Failure to Interrogate	23/09/2022	23/09/2022
							F1901	Additional Surgery	A071102	Failure to Interrogate	23/09/2022	23/09/2022
					CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A071102	Failure to Interrogate	23/09/2022	23/09/2022
							F1901	Additional Surgery	A071102	Failure to Interrogate	23/09/2022	23/09/2022
					CT233	Surgical	E2403	No Clinical Signs, Symptoms or Conditions	A071102	Failure to Interrogate	23/09/2022	23/09/2022
							F1901	Additional Surgery	A071102	Failure to Interrogate	23/09/2022	23/09/2022
					CT2406	Implantable	E2403	No Clinical Signs, Symptoms or Conditions	A071102	Failure to Interrogate	23/09/2022	23/09/2022
							F1901	Additional Surgery	A071102	Failure to Interrogate	23/09/2022	23/09/2022
					CT2407	Active implantable	E2403	No Clinical Signs, Symptoms or Conditions	A071102	Failure to Interrogate	23/09/2022	23/09/2022
							F1901	Additional Surgery	A071102	Failure to Interrogate	23/09/2022	23/09/2022
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A071102	Failure to Interrogate	23/09/2022	23/09/2022
							F1901	Additional Surgery	A071102	Failure to Interrogate	23/09/2022	23/09/2022
					CT2727	Active implant devices/ancillaries	E2403	No Clinical Signs, Symptoms or Conditions	A071102	Failure to Interrogate	23/09/2022	23/09/2022
							F1901	Additional Surgery	A071102	Failure to Interrogate	23/09/2022	23/09/2022
					CT2750	Implantable cardiac electrical stimulators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A071102	Failure to Interrogate	23/09/2022	23/09/2022
							F1901	Additional Surgery	A071102	Failure to Interrogate	23/09/2022	23/09/2022
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A071102	Failure to Interrogate	23/09/2022	23/09/2022
							F1901	Additional Surgery	A071102	Failure to Interrogate	23/09/2022	23/09/2022
					CT2985	Active	E2403	No Clinical Signs, Symptoms or Conditions	A071102	Failure to Interrogate	23/09/2022	23/09/2022
							F1901	Additional Surgery	A071102	Failure to Interrogate	23/09/2022	23/09/2022
					CT3001	Implantable metal	E2403	No Clinical Signs, Symptoms or Conditions	A071102	Failure to Interrogate	23/09/2022	23/09/2022
							F1901	Additional Surgery	A071102	Failure to Interrogate	23/09/2022	23/09/2022
					CT3002	Implantable synthetic polymer	E2403	No Clinical Signs, Symptoms or Conditions	A071102	Failure to Interrogate	23/09/2022	23/09/2022
							F1901	Additional Surgery	A071102	Failure to Interrogate	23/09/2022	23/09/2022
					CT3006	Implantable inorganic material	E2403	No Clinical Signs, Symptoms or Conditions	A071102	Failure to Interrogate	23/09/2022	23/09/2022
							F1901	Additional Surgery	A071102	Failure to Interrogate	23/09/2022	23/09/2022
					CT314	Electrically powered	E2403	No Clinical Signs, Symptoms or Conditions	A071102	Failure to Interrogate	23/09/2022	23/09/2022
							F1901	Additional Surgery	A071102	Failure to Interrogate	23/09/2022	23/09/2022
					CT321	Long-term surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A071102	Failure to Interrogate	23/09/2022	23/09/2022
							F1901	Additional Surgery	A071102	Failure to Interrogate	23/09/2022	23/09/2022
					CT449	Generators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A071102	Failure to Interrogate	23/09/2022	23/09/2022
							F1901	Additional Surgery	A071102	Failure to Interrogate	23/09/2022	23/09/2022
					CT981	Single use	E2403	No Clinical Signs, Symptoms or Conditions	A071102	Failure to Interrogate	23/09/2022	23/09/2022
							F1901	Additional Surgery	A071102	Failure to Interrogate	23/09/2022	23/09/2022
					CT983	Surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A071102	Failure to Interrogate	23/09/2022	23/09/2022
							F1901	Additional Surgery	A071102	Failure to Interrogate	23/09/2022	23/09/2022
7855	2022	All other reportable incidents	47265	Dual-chamber implantable pacemaker, rate-responsive	CT141	Electrodes/Leads and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	22/09/2022	22/09/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	22/09/2022	22/09/2022
					CT145	Cardiovascular devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	22/09/2022	22/09/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	22/09/2022	22/09/2022
					CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	22/09/2022	22/09/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	22/09/2022	22/09/2022
					CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	22/09/2022	22/09/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	22/09/2022	22/09/2022
					CT233	Surgical	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	22/09/2022	22/09/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	22/09/2022	22/09/2022
					CT2406	Implantable	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	22/09/2022	22/09/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	22/09/2022	22/09/2022
					CT2407	Active implantable	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	22/09/2022	22/09/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	22/09/2022	22/09/2022
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	22/09/2022	22/09/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	22/09/2022	22/09/2022
					CT2727	Active implant devices/ancillaries	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	22/09/2022	22/09/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	22/09/2022	22/09/2022
					CT2750	Implantable cardiac electrical stimulators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	22/09/2022	22/09/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	22/09/2022	22/09/2022
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	22/09/2022	22/09/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	22/09/2022	22/09/2022

					CT2985	Active	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	22/09/2022	22/09/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	22/09/2022	22/09/2022
					CT3001	Implantable metal	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	22/09/2022	22/09/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	22/09/2022	22/09/2022
					CT3002	Implantable synthetic polymer	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	22/09/2022	22/09/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	22/09/2022	22/09/2022
					CT3006	Implantable inorganic material	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	22/09/2022	22/09/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	22/09/2022	22/09/2022
					CT314	Electrically powered	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	22/09/2022	22/09/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	22/09/2022	22/09/2022
					CT321	Long-term surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	22/09/2022	22/09/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	22/09/2022	22/09/2022
					CT449	Generators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	22/09/2022	22/09/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	22/09/2022	22/09/2022
					CT981	Single use	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	22/09/2022	22/09/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	22/09/2022	22/09/2022
					CT983	Surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	22/09/2022	22/09/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	22/09/2022	22/09/2022
7856	2022	All other reportable incidents	47265	Dual-chamber implantable pacemaker, rate-responsive	CT141	Electrodes/Leads and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT145	Cardiovascular devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT233	Surgical	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT2406	Implantable	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT2407	Active implantable	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT2727	Active implant devices/ancillaries	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT2750	Implantable cardiac electrical stimulators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT2985	Active	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT3001	Implantable metal	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT3002	Implantable synthetic polymer	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT3006	Implantable inorganic material	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT314	Electrically powered	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT321	Long-term surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT449	Generators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT981	Single use	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT983	Surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
7857	2022	All other reportable incidents	37265	Dual-chamber implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	22/11/2022	22/11/2022
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	22/11/2022	22/11/2022
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	22/11/2022	22/11/2022
7858	2022	All other reportable incidents	47270	Cardiac resynchronization therapy implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
7859	2022	All other reportable incidents	17882	Physiologic-monitoring defibrillation system	CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070912	Incorrect Interpretation of Signal		
							F26	No Health Consequences or Impact	A070912	Incorrect Interpretation of Signal		
7860	2022	All other reportable incidents	17882	Physiologic-monitoring defibrillation system	CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070912	Incorrect Interpretation of Signal		
							F26	No Health Consequences or Impact	A070912	Incorrect Interpretation of Signal		
7861	2022	All other reportable incidents	17882	Physiologic-monitoring defibrillation system	CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070803	Failure to Power Up		
							F27	Problem Identified During Non-Clinical Procedure	A070803	Failure to Power Up		
7862	2022	All other reportable incidents	17882	Physiologic-monitoring defibrillation system	CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070803	Failure to Power Up		
							F27	Problem Identified During Non-Clinical Procedure	A070803	Failure to Power Up		
7863	2022	All other reportable incidents	17882	Physiologic-monitoring defibrillation system	CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A0708	Power Problem		
							F27	Problem Identified During Non-Clinical Procedure	A0708	Power Problem		
7864	2022	All other reportable incidents	60789	Intracardiac pacemaker	CT130	Catheters and associated devices	E011903	Syncope/Fainting	A0701	Capturing Problem		
									A070102	High Capture Threshold		
							E060104	Bradycardia	A0701	Capturing Problem		
									A070102	High Capture Threshold		
							E2403	No Clinical Signs, Symptoms or Conditions	A0701	Capturing Problem		
									A070102	High Capture Threshold		
							F1203	Life Threatening Illness or Injury	A0701	Capturing Problem		
									A070102	High Capture Threshold		
					CT145	Cardiovascular devices	E011903	Syncope/Fainting	A0701	Capturing Problem		
									A070102	High Capture Threshold		
							E060104	Bradycardia	A0701	Capturing Problem		
									A070102	High Capture Threshold		
							E2403	No Clinical Signs, Symptoms or Conditions	A0701	Capturing Problem		

									A070102	High Capture Threshold		
							F1203	Life Threatening Illness or Injury	A0701	Capturing Problem		
									A070102	High Capture Threshold		
					CT1684	Cardiac pacemakers and associated devices	E011903	Syncope/Fainting	A0701	Capturing Problem		
									A070102	High Capture Threshold		
							E060104	Bradycardia	A0701	Capturing Problem		
									A070102	High Capture Threshold		
							E2403	No Clinical Signs, Symptoms or Conditions	A0701	Capturing Problem		
									A070102	High Capture Threshold		
							F1203	Life Threatening Illness or Injury	A0701	Capturing Problem		
									A070102	High Capture Threshold		
					CT1685	Cardiac pacemakers	E011903	Syncope/Fainting	A0701	Capturing Problem		
									A070102	High Capture Threshold		
							E060104	Bradycardia	A0701	Capturing Problem		
									A070102	High Capture Threshold		
							E2403	No Clinical Signs, Symptoms or Conditions	A0701	Capturing Problem		
									A070102	High Capture Threshold		
							F1203	Life Threatening Illness or Injury	A0701	Capturing Problem		
									A070102	High Capture Threshold		
					CT233	Surgical	E011903	Syncope/Fainting	A0701	Capturing Problem		
									A070102	High Capture Threshold		
							E060104	Bradycardia	A0701	Capturing Problem		
									A070102	High Capture Threshold		
							E2403	No Clinical Signs, Symptoms or Conditions	A0701	Capturing Problem		
									A070102	High Capture Threshold		
							F1203	Life Threatening Illness or Injury	A0701	Capturing Problem		
									A070102	High Capture Threshold		
					CT2406	Implantable	E011903	Syncope/Fainting	A0701	Capturing Problem		
									A070102	High Capture Threshold		
							E060104	Bradycardia	A0701	Capturing Problem		
									A070102	High Capture Threshold		
							E2403	No Clinical Signs, Symptoms or Conditions	A0701	Capturing Problem		
									A070102	High Capture Threshold		
							F1203	Life Threatening Illness or Injury	A0701	Capturing Problem		
									A070102	High Capture Threshold		
					CT2407	Active implantable	E011903	Syncope/Fainting	A0701	Capturing Problem		
									A070102	High Capture Threshold		
							E060104	Bradycardia	A0701	Capturing Problem		
									A070102	High Capture Threshold		
							E2403	No Clinical Signs, Symptoms or Conditions	A0701	Capturing Problem		
									A070102	High Capture Threshold		
							F1203	Life Threatening Illness or Injury	A0701	Capturing Problem		
									A070102	High Capture Threshold		
					CT257	Implantable cardiac pacemakers	E011903	Syncope/Fainting	A0701	Capturing Problem		
									A070102	High Capture Threshold		
							E060104	Bradycardia	A0701	Capturing Problem		
									A070102	High Capture Threshold		
							E2403	No Clinical Signs, Symptoms or Conditions	A0701	Capturing Problem		
									A070102	High Capture Threshold		
							F1203	Life Threatening Illness or Injury	A0701	Capturing Problem		
									A070102	High Capture Threshold		
					CT2727	Active implant devices/ancillaries	E011903	Syncope/Fainting	A0701	Capturing Problem		
									A070102	High Capture Threshold		
							E060104	Bradycardia	A0701	Capturing Problem		
									A070102	High Capture Threshold		
							E2403	No Clinical Signs, Symptoms or Conditions	A0701	Capturing Problem		
									A070102	High Capture Threshold		
							F1203	Life Threatening Illness or Injury	A0701	Capturing Problem		
									A070102	High Capture Threshold		
					CT2750	Implantable cardiac electrical stimulators and associated devices	E011903	Syncope/Fainting	A0701	Capturing Problem		
									A070102	High Capture Threshold		
							E060104	Bradycardia	A0701	Capturing Problem		
									A070102	High Capture Threshold		
							E2403	No Clinical Signs, Symptoms or Conditions	A0701	Capturing Problem		
									A070102	High Capture Threshold		
							F1203	Life Threatening Illness or Injury	A0701	Capturing Problem		
									A070102	High Capture Threshold		
					CT2818	Catheterization/Cannulation/Fluid path devices	E011903	Syncope/Fainting	A0701	Capturing Problem		
									A070102	High Capture Threshold		
							E060104	Bradycardia	A0701	Capturing Problem		
									A070102	High Capture Threshold		
							E2403	No Clinical Signs, Symptoms or Conditions	A0701	Capturing Problem		
									A070102	High Capture Threshold		
							F1203	Life Threatening Illness or Injury	A0701	Capturing Problem		
									A070102	High Capture Threshold		
					CT2866	Implantable cardiac electrical stimulators	E011903	Syncope/Fainting	A0701	Capturing Problem		
									A070102	High Capture Threshold		
							E060104	Bradycardia	A0701	Capturing Problem		
									A070102	High Capture Threshold		
							E2403	No Clinical Signs, Symptoms or Conditions	A0701	Capturing Problem		
									A070102	High Capture Threshold		
							F1203	Life Threatening Illness or Injury	A0701	Capturing Problem		
									A070102	High Capture Threshold		
					CT2985	Active	E011903	Syncope/Fainting	A0701	Capturing Problem		
									A070102	High Capture Threshold		
							E060104	Bradycardia	A0701	Capturing Problem		
									A070102	High Capture Threshold		
							E2403	No Clinical Signs, Symptoms or Conditions	A0701	Capturing Problem		

									A070102	High Capture Threshold		
							F1203	Life Threatening Illness or Injury	A0701	Capturing Problem		
									A070102	High Capture Threshold		
					CT3001	Implantable metal	E011903	Syncope/Fainting	A0701	Capturing Problem		
									A070102	High Capture Threshold		
							E060104	Bradycardia	A0701	Capturing Problem		
									A070102	High Capture Threshold		
							E2403	No Clinical Signs, Symptoms or Conditions	A0701	Capturing Problem		
									A070102	High Capture Threshold		
							F1203	Life Threatening Illness or Injury	A0701	Capturing Problem		
									A070102	High Capture Threshold		
					CT3002	Implantable synthetic polymer	E011903	Syncope/Fainting	A0701	Capturing Problem		
									A070102	High Capture Threshold		
							E060104	Bradycardia	A0701	Capturing Problem		
									A070102	High Capture Threshold		
							E2403	No Clinical Signs, Symptoms or Conditions	A0701	Capturing Problem		
									A070102	High Capture Threshold		
							F1203	Life Threatening Illness or Injury	A0701	Capturing Problem		
									A070102	High Capture Threshold		
					CT3006	Implantable inorganic material	E011903	Syncope/Fainting	A0701	Capturing Problem		
									A070102	High Capture Threshold		
							E060104	Bradycardia	A0701	Capturing Problem		
									A070102	High Capture Threshold		
							E2403	No Clinical Signs, Symptoms or Conditions	A0701	Capturing Problem		
									A070102	High Capture Threshold		
							F1203	Life Threatening Illness or Injury	A0701	Capturing Problem		
									A070102	High Capture Threshold		
					CT314	Electrically powered	E011903	Syncope/Fainting	A0701	Capturing Problem		
									A070102	High Capture Threshold		
							E060104	Bradycardia	A0701	Capturing Problem		
									A070102	High Capture Threshold		
							E2403	No Clinical Signs, Symptoms or Conditions	A0701	Capturing Problem		
									A070102	High Capture Threshold		
							F1203	Life Threatening Illness or Injury	A0701	Capturing Problem		
									A070102	High Capture Threshold		
					CT321	Long-term surgical invasive	E011903	Syncope/Fainting	A0701	Capturing Problem		
									A070102	High Capture Threshold		
							E060104	Bradycardia	A0701	Capturing Problem		
									A070102	High Capture Threshold		
							E2403	No Clinical Signs, Symptoms or Conditions	A0701	Capturing Problem		
									A070102	High Capture Threshold		
							F1203	Life Threatening Illness or Injury	A0701	Capturing Problem		
									A070102	High Capture Threshold		
					CT981	Single use	E011903	Syncope/Fainting	A0701	Capturing Problem		
									A070102	High Capture Threshold		
							E060104	Bradycardia	A0701	Capturing Problem		
									A070102	High Capture Threshold		
							E2403	No Clinical Signs, Symptoms or Conditions	A0701	Capturing Problem		
									A070102	High Capture Threshold		
							F1203	Life Threatening Illness or Injury	A0701	Capturing Problem		
									A070102	High Capture Threshold		
					CT983	Surgical invasive	E011903	Syncope/Fainting	A0701	Capturing Problem		
									A070102	High Capture Threshold		
							E060104	Bradycardia	A0701	Capturing Problem		
									A070102	High Capture Threshold		
							E2403	No Clinical Signs, Symptoms or Conditions	A0701	Capturing Problem		
									A070102	High Capture Threshold		
							F1203	Life Threatening Illness or Injury	A0701	Capturing Problem		
									A070102	High Capture Threshold		
7865	2022	All other reportable incidents	47265	Dual-chamber implantable pacemaker, rate-responsive	CT141	Electrodes/Leads and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	10/05/2022	10/05/2022
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	10/05/2022	10/05/2022
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	10/05/2022	10/05/2022
					CT145	Cardiovascular devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	10/05/2022	10/05/2022
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	10/05/2022	10/05/2022
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	10/05/2022	10/05/2022
					CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	10/05/2022	10/05/2022
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	10/05/2022	10/05/2022
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	10/05/2022	10/05/2022
					CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	10/05/2022	10/05/2022
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	10/05/2022	10/05/2022
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	10/05/2022	10/05/2022
					CT233	Surgical	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	10/05/2022	10/05/2022
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	10/05/2022	10/05/2022
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	10/05/2022	10/05/2022
					CT2406	Implantable	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	10/05/2022	10/05/2022
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	10/05/2022	10/05/2022
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	10/05/2022	10/05/2022
					CT2407	Active implantable	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	10/05/2022	10/05/2022
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	10/05/2022	10/05/2022
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	10/05/2022	10/05/2022
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	10/05/2022	10/05/2022
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	10/05/2022	10/05/2022
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	10/05/2022	10/05/2022
					CT2727	Active implant devices/ancillaries	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	10/05/2022	10/05/2022
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	10/05/2022	10/05/2022
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	10/05/2022	10/05/2022
					CT2750	Implantable cardiac electrical stimulators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	10/05/2022	10/05/2022
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	10/05/2022	10/05/2022

							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	10/05/2022	10/05/2022
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	10/05/2022	10/05/2022
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	10/05/2022	10/05/2022
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	10/05/2022	10/05/2022
					CT2985	Active	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	10/05/2022	10/05/2022
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	10/05/2022	10/05/2022
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	10/05/2022	10/05/2022
					CT3001	Implantable metal	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	10/05/2022	10/05/2022
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	10/05/2022	10/05/2022
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	10/05/2022	10/05/2022
					CT3002	Implantable synthetic polymer	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	10/05/2022	10/05/2022
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	10/05/2022	10/05/2022
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	10/05/2022	10/05/2022
					CT3006	Implantable inorganic material	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	10/05/2022	10/05/2022
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	10/05/2022	10/05/2022
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	10/05/2022	10/05/2022
					CT314	Electrically powered	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	10/05/2022	10/05/2022
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	10/05/2022	10/05/2022
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	10/05/2022	10/05/2022
					CT321	Long-term surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	10/05/2022	10/05/2022
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	10/05/2022	10/05/2022
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	10/05/2022	10/05/2022
					CT449	Generators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	10/05/2022	10/05/2022
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	10/05/2022	10/05/2022
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	10/05/2022	10/05/2022
					CT981	Single use	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	10/05/2022	10/05/2022
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	10/05/2022	10/05/2022
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	10/05/2022	10/05/2022
					CT983	Surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	10/05/2022	10/05/2022
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	10/05/2022	10/05/2022
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	10/05/2022	10/05/2022
7866	2022	All other reportable incidents	47265	Dual-chamber implantable pacemaker, rate-responsive	CT141	Electrodes/Leads and associated devices	E0717	Dyspnea	A070504	Premature Discharge of Battery	27/01/2022	27/01/2022
							E2312	Fatigue	A070504	Premature Discharge of Battery	27/01/2022	27/01/2022
							E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	27/01/2022	27/01/2022
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	27/01/2022	27/01/2022
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	27/01/2022	27/01/2022
					CT145	Cardiovascular devices	E0717	Dyspnea	A070504	Premature Discharge of Battery	27/01/2022	27/01/2022
							E2312	Fatigue	A070504	Premature Discharge of Battery	27/01/2022	27/01/2022
							E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	27/01/2022	27/01/2022
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	27/01/2022	27/01/2022
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	27/01/2022	27/01/2022
					CT1684	Cardiac pacemakers and associated devices	E0717	Dyspnea	A070504	Premature Discharge of Battery	27/01/2022	27/01/2022
							E2312	Fatigue	A070504	Premature Discharge of Battery	27/01/2022	27/01/2022
							E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	27/01/2022	27/01/2022
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	27/01/2022	27/01/2022
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	27/01/2022	27/01/2022
					CT1685	Cardiac pacemakers	E0717	Dyspnea	A070504	Premature Discharge of Battery	27/01/2022	27/01/2022
							E2312	Fatigue	A070504	Premature Discharge of Battery	27/01/2022	27/01/2022
							E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	27/01/2022	27/01/2022
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	27/01/2022	27/01/2022
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	27/01/2022	27/01/2022
					CT233	Surgical	E0717	Dyspnea	A070504	Premature Discharge of Battery	27/01/2022	27/01/2022
							E2312	Fatigue	A070504	Premature Discharge of Battery	27/01/2022	27/01/2022
							E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	27/01/2022	27/01/2022
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	27/01/2022	27/01/2022
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	27/01/2022	27/01/2022
					CT2406	Implantable	E0717	Dyspnea	A070504	Premature Discharge of Battery	27/01/2022	27/01/2022
							E2312	Fatigue	A070504	Premature Discharge of Battery	27/01/2022	27/01/2022
							E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	27/01/2022	27/01/2022
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	27/01/2022	27/01/2022
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	27/01/2022	27/01/2022
					CT2407	Active implantable	E0717	Dyspnea	A070504	Premature Discharge of Battery	27/01/2022	27/01/2022
							E2312	Fatigue	A070504	Premature Discharge of Battery	27/01/2022	27/01/2022
							E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	27/01/2022	27/01/2022
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	27/01/2022	27/01/2022
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	27/01/2022	27/01/2022
					CT257	Implantable cardiac pacemakers	E0717	Dyspnea	A070504	Premature Discharge of Battery	27/01/2022	27/01/2022
							E2312	Fatigue	A070504	Premature Discharge of Battery	27/01/2022	27/01/2022
							E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	27/01/2022	27/01/2022
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	27/01/2022	27/01/2022
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	27/01/2022	27/01/2022
					CT2727	Active implant devices/ancillaries	E0717	Dyspnea	A070504	Premature Discharge of Battery	27/01/2022	27/01/2022
							E2312	Fatigue	A070504	Premature Discharge of Battery	27/01/2022	27/01/2022
							E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	27/01/2022	27/01/2022
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	27/01/2022	27/01/2022
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	27/01/2022	27/01/2022
					CT2750	Implantable cardiac electrical stimulators and associated devices	E0717	Dyspnea	A070504	Premature Discharge of Battery	27/01/2022	27/01/2022
							E2312	Fatigue	A070504	Premature Discharge of Battery	27/01/2022	27/01/2022
							E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	27/01/2022	27/01/2022
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	27/01/2022	27/01/2022
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	27/01/2022	27/01/2022
					CT2866	Implantable cardiac electrical stimulators	E0717	Dyspnea	A070504	Premature Discharge of Battery	27/01/2022	27/01/2022
							E2312	Fatigue	A070504	Premature Discharge of Battery	27/01/2022	27/01/2022
							E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	27/01/2022	27/01/2022
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	27/01/2022	27/01/2022
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	27/01/2022	27/01/2022
					CT2985	Active	E0717	Dyspnea	A070504	Premature Discharge of Battery	27/01/2022	27/01/2022
							E2312	Fatigue	A070504	Premature Discharge of Battery	27/01/2022	27/01/2022

							E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	27/01/2022	27/01/2022
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	27/01/2022	27/01/2022
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	27/01/2022	27/01/2022
					CT3001	Implantable metal	E0717	Dyspnea	A070504	Premature Discharge of Battery	27/01/2022	27/01/2022
							E2312	Fatigue	A070504	Premature Discharge of Battery	27/01/2022	27/01/2022
							E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	27/01/2022	27/01/2022
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	27/01/2022	27/01/2022
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	27/01/2022	27/01/2022
					CT3002	Implantable synthetic polymer	E0717	Dyspnea	A070504	Premature Discharge of Battery	27/01/2022	27/01/2022
							E2312	Fatigue	A070504	Premature Discharge of Battery	27/01/2022	27/01/2022
							E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	27/01/2022	27/01/2022
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	27/01/2022	27/01/2022
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	27/01/2022	27/01/2022
					CT3006	Implantable inorganic material	E0717	Dyspnea	A070504	Premature Discharge of Battery	27/01/2022	27/01/2022
							E2312	Fatigue	A070504	Premature Discharge of Battery	27/01/2022	27/01/2022
							E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	27/01/2022	27/01/2022
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	27/01/2022	27/01/2022
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	27/01/2022	27/01/2022
					CT314	Electrically powered	E0717	Dyspnea	A070504	Premature Discharge of Battery	27/01/2022	27/01/2022
							E2312	Fatigue	A070504	Premature Discharge of Battery	27/01/2022	27/01/2022
							E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	27/01/2022	27/01/2022
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	27/01/2022	27/01/2022
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	27/01/2022	27/01/2022
					CT321	Long-term surgical invasive	E0717	Dyspnea	A070504	Premature Discharge of Battery	27/01/2022	27/01/2022
							E2312	Fatigue	A070504	Premature Discharge of Battery	27/01/2022	27/01/2022
							E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	27/01/2022	27/01/2022
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	27/01/2022	27/01/2022
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	27/01/2022	27/01/2022
					CT449	Generators and associated devices	E0717	Dyspnea	A070504	Premature Discharge of Battery	27/01/2022	27/01/2022
							E2312	Fatigue	A070504	Premature Discharge of Battery	27/01/2022	27/01/2022
							E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	27/01/2022	27/01/2022
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	27/01/2022	27/01/2022
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	27/01/2022	27/01/2022
					CT981	Single use	E0717	Dyspnea	A070504	Premature Discharge of Battery	27/01/2022	27/01/2022
							E2312	Fatigue	A070504	Premature Discharge of Battery	27/01/2022	27/01/2022
							E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	27/01/2022	27/01/2022
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	27/01/2022	27/01/2022
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	27/01/2022	27/01/2022
					CT983	Surgical invasive	E0717	Dyspnea	A070504	Premature Discharge of Battery	27/01/2022	27/01/2022
							E2312	Fatigue	A070504	Premature Discharge of Battery	27/01/2022	27/01/2022
							E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	27/01/2022	27/01/2022
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	27/01/2022	27/01/2022
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	27/01/2022	27/01/2022
7867	2022	All other reportable incidents	17882	Physiologic-monitoring defibrillation system	CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A07	Electrical /Electronic Property Problem		
							F27	Problem Identified During Non-Clinical Procedure	A07	Electrical /Electronic Property Problem		
7868	2022	All other reportable incidents	36241	Pacemaker/defibrillator lead	CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A040101	Fracture		
							F26	No Health Consequences or Impact	A040101	Fracture		
7869	2022	All other reportable incidents	35852	Single-chamber implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
7870	2022	All other reportable incidents	35852	Single-chamber implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E2104	Electric Shock	A040101	Fracture		
									A070909	Oversensing		
									A071302	Inappropriate/Inadequate Shock/Stimulation		
							E2403	No Clinical Signs, Symptoms or Conditions	A040101	Fracture		
									A070909	Oversensing		
									A071302	Inappropriate/Inadequate Shock/Stimulation		
							F23	Unexpected Medical Intervention	A040101	Fracture		
									A070909	Oversensing		
									A071302	Inappropriate/Inadequate Shock/Stimulation		
7871	2022	All other reportable incidents	37265	Dual-chamber implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A0508	Noise, Audible	25/09/2022	25/09/2022
									A070504	Premature Discharge of Battery	25/09/2022	25/09/2022
							F08	Hospitalization or Prolonged Hospitalization	A0508	Noise, Audible	25/09/2022	25/09/2022
									A070504	Premature Discharge of Battery	25/09/2022	25/09/2022
							F1905	Device Revision or Replacement	A0508	Noise, Audible	25/09/2022	25/09/2022
									A070504	Premature Discharge of Battery	25/09/2022	25/09/2022
7872	2022	All other reportable incidents	17882	Physiologic-monitoring defibrillation system	CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070803	Failure to Power Up		
							F27	Problem Identified During Non-Clinical Procedure	A070803	Failure to Power Up		
7873	2022	All other reportable incidents	17882	Physiologic-monitoring defibrillation system	CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070803	Failure to Power Up		
							F27	Problem Identified During Non-Clinical Procedure	A070803	Failure to Power Up		
7874	2022	All other reportable incidents	17882	Physiologic-monitoring defibrillation system	CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070803	Failure to Power Up		
							F27	Problem Identified During Non-Clinical Procedure	A070803	Failure to Power Up		
7875	2022	All other reportable incidents	47265	Dual-chamber implantable pacemaker, rate-responsive	CT141	Electrodes/Leads and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
					CT145	Cardiovascular devices	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
					CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
					CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
					CT233	Surgical	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		

							F26	No Health Consequences or Impact	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
					CT2406	Implantable	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
					CT2407	Active implantable	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
					CT2727	Active implant devices/ancillaries	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
					CT2750	Implantable cardiac electrical stimulators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
					CT2985	Active	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
					CT3001	Implantable metal	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
					CT3002	Implantable synthetic polymer	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
					CT3006	Implantable inorganic material	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
					CT314	Electrically powered	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
					CT321	Long-term surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
					CT449	Generators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
					CT981	Single use	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
					CT983	Surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
7876	2022	All other reportable incidents	47263	Cardiac resynchronization therapy implantable pacemaker	CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A0908	Incorrect, Inadequate or Imprecise Result or Readings	08/02/2022	08/02/2022
							F08	Hospitalization or Prolonged Hospitalization	A0908	Incorrect, Inadequate or Imprecise Result or Readings	08/02/2022	08/02/2022
							F1905	Device Revision or Replacement	A0908	Incorrect, Inadequate or Imprecise Result or Readings	08/02/2022	08/02/2022
					CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A0908	Incorrect, Inadequate or Imprecise Result or Readings	08/02/2022	08/02/2022
							F08	Hospitalization or Prolonged Hospitalization	A0908	Incorrect, Inadequate or Imprecise Result or Readings	08/02/2022	08/02/2022
							F1905	Device Revision or Replacement	A0908	Incorrect, Inadequate or Imprecise Result or Readings	08/02/2022	08/02/2022
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A0908	Incorrect, Inadequate or Imprecise Result or Readings	08/02/2022	08/02/2022
							F08	Hospitalization or Prolonged Hospitalization	A0908	Incorrect, Inadequate or Imprecise Result or Readings	08/02/2022	08/02/2022
							F1905	Device Revision or Replacement	A0908	Incorrect, Inadequate or Imprecise Result or Readings	08/02/2022	08/02/2022
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A0908	Incorrect, Inadequate or Imprecise Result or Readings	08/02/2022	08/02/2022
							F08	Hospitalization or Prolonged Hospitalization	A0908	Incorrect, Inadequate or Imprecise Result or Readings	08/02/2022	08/02/2022
							F1905	Device Revision or Replacement	A0908	Incorrect, Inadequate or Imprecise Result or Readings	08/02/2022	08/02/2022
7877	2022	All other reportable incidents	35852	Single-chamber implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E2104	Electric Shock	A070909	Oversensing		
									A071302	Inappropriate/Inadequate Shock/Stimulation		
									A090801	Signal Artifact/Noise		
							E2403	No Clinical Signs, Symptoms or Conditions	A070909	Oversensing		
									A071302	Inappropriate/Inadequate Shock/Stimulation		

									A090801	Signal Artifact/Noise		
						F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A070909	Oversensing			
								A071302	Inappropriate/Inadequate Shock/Stimulation			
								A090801	Signal Artifact/Noise			
7878	2022	All other reportable incidents	35852	Single-chamber implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	19/12/2020	19/12/2020
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	19/12/2020	19/12/2020
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	19/12/2020	19/12/2020
7879	2022	All other reportable incidents	47265	Dual-chamber implantable pacemaker, rate-responsive	CT141	Electrodes/Leads and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A072201	High Impedance		
							F26	No Health Consequences or Impact	A072201	High Impedance		
					CT145	Cardiovascular devices	E2403	No Clinical Signs, Symptoms or Conditions	A072201	High Impedance		
							F26	No Health Consequences or Impact	A072201	High Impedance		
					CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A072201	High Impedance		
							F26	No Health Consequences or Impact	A072201	High Impedance		
					CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A072201	High Impedance		
							F26	No Health Consequences or Impact	A072201	High Impedance		
					CT233	Surgical	E2403	No Clinical Signs, Symptoms or Conditions	A072201	High Impedance		
							F26	No Health Consequences or Impact	A072201	High Impedance		
					CT2406	Implantable	E2403	No Clinical Signs, Symptoms or Conditions	A072201	High Impedance		
							F26	No Health Consequences or Impact	A072201	High Impedance		
					CT2407	Active implantable	E2403	No Clinical Signs, Symptoms or Conditions	A072201	High Impedance		
							F26	No Health Consequences or Impact	A072201	High Impedance		
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A072201	High Impedance		
							F26	No Health Consequences or Impact	A072201	High Impedance		
					CT2727	Active implant devices/ancillaries	E2403	No Clinical Signs, Symptoms or Conditions	A072201	High Impedance		
							F26	No Health Consequences or Impact	A072201	High Impedance		
					CT2750	Implantable cardiac electrical stimulators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A072201	High Impedance		
							F26	No Health Consequences or Impact	A072201	High Impedance		
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A072201	High Impedance		
							F26	No Health Consequences or Impact	A072201	High Impedance		
					CT2985	Active	E2403	No Clinical Signs, Symptoms or Conditions	A072201	High Impedance		
							F26	No Health Consequences or Impact	A072201	High Impedance		
					CT3001	Implantable metal	E2403	No Clinical Signs, Symptoms or Conditions	A072201	High Impedance		
							F26	No Health Consequences or Impact	A072201	High Impedance		
					CT3002	Implantable synthetic polymer	E2403	No Clinical Signs, Symptoms or Conditions	A072201	High Impedance		
							F26	No Health Consequences or Impact	A072201	High Impedance		
					CT3006	Implantable inorganic material	E2403	No Clinical Signs, Symptoms or Conditions	A072201	High Impedance		
							F26	No Health Consequences or Impact	A072201	High Impedance		
					CT314	Electrically powered	E2403	No Clinical Signs, Symptoms or Conditions	A072201	High Impedance		
							F26	No Health Consequences or Impact	A072201	High Impedance		
					CT321	Long-term surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A072201	High Impedance		
							F26	No Health Consequences or Impact	A072201	High Impedance		
					CT449	Generators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A072201	High Impedance		
							F26	No Health Consequences or Impact	A072201	High Impedance		
					CT981	Single use	E2403	No Clinical Signs, Symptoms or Conditions	A072201	High Impedance		
							F26	No Health Consequences or Impact	A072201	High Impedance		
					CT983	Surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A072201	High Impedance		
							F26	No Health Consequences or Impact	A072201	High Impedance		
7880	2022	All other reportable incidents	36241	Pacemaker/defibrillator lead	CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A072201	High Impedance		
							F26	No Health Consequences or Impact	A072201	High Impedance		
7881	2022	All other reportable incidents	47267	Single-chamber implantable pacemaker, rate-responsive	CT141	Electrodes/Leads and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A072201	High Impedance		
							F26	No Health Consequences or Impact	A072201	High Impedance		
					CT145	Cardiovascular devices	E2403	No Clinical Signs, Symptoms or Conditions	A072201	High Impedance		
							F26	No Health Consequences or Impact	A072201	High Impedance		
					CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A072201	High Impedance		
							F26	No Health Consequences or Impact	A072201	High Impedance		
					CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A072201	High Impedance		
							F26	No Health Consequences or Impact	A072201	High Impedance		
					CT233	Surgical	E2403	No Clinical Signs, Symptoms or Conditions	A072201	High Impedance		
							F26	No Health Consequences or Impact	A072201	High Impedance		
					CT2406	Implantable	E2403	No Clinical Signs, Symptoms or Conditions	A072201	High Impedance		
							F26	No Health Consequences or Impact	A072201	High Impedance		
					CT2407	Active implantable	E2403	No Clinical Signs, Symptoms or Conditions	A072201	High Impedance		
							F26	No Health Consequences or Impact	A072201	High Impedance		
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A072201	High Impedance		
							F26	No Health Consequences or Impact	A072201	High Impedance		
					CT2727	Active implant devices/ancillaries	E2403	No Clinical Signs, Symptoms or Conditions	A072201	High Impedance		
							F26	No Health Consequences or Impact	A072201	High Impedance		
					CT2750	Implantable cardiac electrical stimulators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A072201	High Impedance		
							F26	No Health Consequences or Impact	A072201	High Impedance		
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A072201	High Impedance		
							F26	No Health Consequences or Impact	A072201	High Impedance		
					CT2985	Active	E2403	No Clinical Signs, Symptoms or Conditions	A072201	High Impedance		
							F26	No Health Consequences or Impact	A072201	High Impedance		
					CT3001	Implantable metal	E2403	No Clinical Signs, Symptoms or Conditions	A072201	High Impedance		
							F26	No Health Consequences or Impact	A072201	High Impedance		
					CT3002	Implantable synthetic polymer	E2403	No Clinical Signs, Symptoms or Conditions	A072201	High Impedance		
							F26	No Health Consequences or Impact	A072201	High Impedance		
					CT3006	Implantable inorganic material	E2403	No Clinical Signs, Symptoms or Conditions	A072201	High Impedance		
							F26	No Health Consequences or Impact	A072201	High Impedance		
					CT314	Electrically powered	E2403	No Clinical Signs, Symptoms or Conditions	A072201	High Impedance		
							F26	No Health Consequences or Impact	A072201	High Impedance		
					CT321	Long-term surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A072201	High Impedance		
							F26	No Health Consequences or Impact	A072201	High Impedance		
					CT449	Generators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A072201	High Impedance		
							F26	No Health Consequences or Impact	A072201	High Impedance		
					CT981	Single use	E2403	No Clinical Signs, Symptoms or Conditions	A072201	High Impedance		
							F26	No Health Consequences or Impact	A072201	High Impedance		
					CT983	Surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A072201	High Impedance		

							F26	No Health Consequences or Impact	A072201	High Impedance			
7882	2022	All other reportable incidents	47267	Single-chamber implantable pacemaker, rate-responsive	CT141	Electrodes/Leads and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem	13/06/2022	13/06/2022
									A070504	Premature Discharge of Battery	13/06/2022	13/06/2022
							F08	Hospitalization or Prolonged Hospitalization	A0705	Battery Problem	13/06/2022	13/06/2022
									A070504	Premature Discharge of Battery	13/06/2022	13/06/2022
							F1905	Device Revision or Replacement	A0705	Battery Problem	13/06/2022	13/06/2022
									A070504	Premature Discharge of Battery	13/06/2022	13/06/2022
					CT145	Cardiovascular devices	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem	13/06/2022	13/06/2022
									A070504	Premature Discharge of Battery	13/06/2022	13/06/2022
							F08	Hospitalization or Prolonged Hospitalization	A0705	Battery Problem	13/06/2022	13/06/2022
									A070504	Premature Discharge of Battery	13/06/2022	13/06/2022
							F1905	Device Revision or Replacement	A0705	Battery Problem	13/06/2022	13/06/2022
									A070504	Premature Discharge of Battery	13/06/2022	13/06/2022
					CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem	13/06/2022	13/06/2022
									A070504	Premature Discharge of Battery	13/06/2022	13/06/2022
							F08	Hospitalization or Prolonged Hospitalization	A0705	Battery Problem	13/06/2022	13/06/2022
									A070504	Premature Discharge of Battery	13/06/2022	13/06/2022
							F1905	Device Revision or Replacement	A0705	Battery Problem	13/06/2022	13/06/2022
									A070504	Premature Discharge of Battery	13/06/2022	13/06/2022
					CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem	13/06/2022	13/06/2022
									A070504	Premature Discharge of Battery	13/06/2022	13/06/2022
							F08	Hospitalization or Prolonged Hospitalization	A0705	Battery Problem	13/06/2022	13/06/2022
									A070504	Premature Discharge of Battery	13/06/2022	13/06/2022
							F1905	Device Revision or Replacement	A0705	Battery Problem	13/06/2022	13/06/2022
									A070504	Premature Discharge of Battery	13/06/2022	13/06/2022
					CT233	Surgical	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem	13/06/2022	13/06/2022
									A070504	Premature Discharge of Battery	13/06/2022	13/06/2022
							F08	Hospitalization or Prolonged Hospitalization	A0705	Battery Problem	13/06/2022	13/06/2022
									A070504	Premature Discharge of Battery	13/06/2022	13/06/2022
							F1905	Device Revision or Replacement	A0705	Battery Problem	13/06/2022	13/06/2022
									A070504	Premature Discharge of Battery	13/06/2022	13/06/2022
					CT2406	Implantable	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem	13/06/2022	13/06/2022
									A070504	Premature Discharge of Battery	13/06/2022	13/06/2022
							F08	Hospitalization or Prolonged Hospitalization	A0705	Battery Problem	13/06/2022	13/06/2022
									A070504	Premature Discharge of Battery	13/06/2022	13/06/2022
							F1905	Device Revision or Replacement	A0705	Battery Problem	13/06/2022	13/06/2022
									A070504	Premature Discharge of Battery	13/06/2022	13/06/2022
					CT2407	Active implantable	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem	13/06/2022	13/06/2022
									A070504	Premature Discharge of Battery	13/06/2022	13/06/2022
							F08	Hospitalization or Prolonged Hospitalization	A0705	Battery Problem	13/06/2022	13/06/2022
									A070504	Premature Discharge of Battery	13/06/2022	13/06/2022
							F1905	Device Revision or Replacement	A0705	Battery Problem	13/06/2022	13/06/2022
									A070504	Premature Discharge of Battery	13/06/2022	13/06/2022
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem	13/06/2022	13/06/2022
									A070504	Premature Discharge of Battery	13/06/2022	13/06/2022
							F08	Hospitalization or Prolonged Hospitalization	A0705	Battery Problem	13/06/2022	13/06/2022
									A070504	Premature Discharge of Battery	13/06/2022	13/06/2022
							F1905	Device Revision or Replacement	A0705	Battery Problem	13/06/2022	13/06/2022
									A070504	Premature Discharge of Battery	13/06/2022	13/06/2022
					CT2727	Active implant devices/ancillaries	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem	13/06/2022	13/06/2022
									A070504	Premature Discharge of Battery	13/06/2022	13/06/2022
							F08	Hospitalization or Prolonged Hospitalization	A0705	Battery Problem	13/06/2022	13/06/2022
									A070504	Premature Discharge of Battery	13/06/2022	13/06/2022
							F1905	Device Revision or Replacement	A0705	Battery Problem	13/06/2022	13/06/2022
									A070504	Premature Discharge of Battery	13/06/2022	13/06/2022
					CT2750	Implantable cardiac electrical stimulators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem	13/06/2022	13/06/2022
									A070504	Premature Discharge of Battery	13/06/2022	13/06/2022
							F08	Hospitalization or Prolonged Hospitalization	A0705	Battery Problem	13/06/2022	13/06/2022
									A070504	Premature Discharge of Battery	13/06/2022	13/06/2022
							F1905	Device Revision or Replacement	A0705	Battery Problem	13/06/2022	13/06/2022
									A070504	Premature Discharge of Battery	13/06/2022	13/06/2022
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem	13/06/2022	13/06/2022
									A070504	Premature Discharge of Battery	13/06/2022	13/06/2022
							F08	Hospitalization or Prolonged Hospitalization	A0705	Battery Problem	13/06/2022	13/06/2022
									A070504	Premature Discharge of Battery	13/06/2022	13/06/2022
							F1905	Device Revision or Replacement	A0705	Battery Problem	13/06/2022	13/06/2022
									A070504	Premature Discharge of Battery	13/06/2022	13/06/2022
					CT2985	Active	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem	13/06/2022	13/06/2022
									A070504	Premature Discharge of Battery	13/06/2022	13/06/2022
							F08	Hospitalization or Prolonged Hospitalization	A0705	Battery Problem	13/06/2022	13/06/2022
									A070504	Premature Discharge of Battery	13/06/2022	13/06/2022
							F1905	Device Revision or Replacement	A0705	Battery Problem	13/06/2022	13/06/2022
									A070504	Premature Discharge of Battery	13/06/2022	13/06/2022
					CT3001	Implantable metal	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem	13/06/2022	13/06/2022
									A070504	Premature Discharge of Battery	13/06/2022	13/06/2022
							F08	Hospitalization or Prolonged Hospitalization	A0705	Battery Problem	13/06/2022	13/06/2022
									A070504	Premature Discharge of Battery	13/06/2022	13/06/2022
							F1905	Device Revision or Replacement	A0705	Battery Problem	13/06/2022	13/06/2022
									A070504	Premature Discharge of Battery	13/06/2022	13/06/2022
					CT3002	Implantable synthetic polymer	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem	13/06/2022	13/06/2022
									A070504	Premature Discharge of Battery	13/06/2022	13/06/2022
							F08	Hospitalization or Prolonged Hospitalization	A0705	Battery Problem	13/06/2022	13/06/2022
									A070504	Premature Discharge of Battery	13/06/2022	13/06/2022
							F1905	Device Revision or Replacement	A0705	Battery Problem	13/06/2022	13/06/2022
									A070504	Premature Discharge of Battery	13/06/2022	13/06/2022
					CT3006	Implantable inorganic material	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem	13/06/2022	13/06/2022
									A070504	Premature Discharge of Battery	13/06/2022	13/06/2022
							F08	Hospitalization or Prolonged Hospitalization	A0705	Battery Problem	13/06/2022	13/06/2022

								A070504	Premature Discharge of Battery	13/06/2022	13/06/2022
						F1905	Device Revision or Replacement	A0705	Battery Problem	13/06/2022	13/06/2022
								A070504	Premature Discharge of Battery	13/06/2022	13/06/2022
					CT314	Electrically powered	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem	13/06/2022	13/06/2022
									A070504	Premature Discharge of Battery	13/06/2022	13/06/2022
							F08	Hospitalization or Prolonged Hospitalization	A0705	Battery Problem	13/06/2022	13/06/2022
									A070504	Premature Discharge of Battery	13/06/2022	13/06/2022
							F1905	Device Revision or Replacement	A0705	Battery Problem	13/06/2022	13/06/2022
									A070504	Premature Discharge of Battery	13/06/2022	13/06/2022
					CT321	Long-term surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem	13/06/2022	13/06/2022
									A070504	Premature Discharge of Battery	13/06/2022	13/06/2022
							F08	Hospitalization or Prolonged Hospitalization	A0705	Battery Problem	13/06/2022	13/06/2022
									A070504	Premature Discharge of Battery	13/06/2022	13/06/2022
							F1905	Device Revision or Replacement	A0705	Battery Problem	13/06/2022	13/06/2022
									A070504	Premature Discharge of Battery	13/06/2022	13/06/2022
					CT449	Generators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem	13/06/2022	13/06/2022
									A070504	Premature Discharge of Battery	13/06/2022	13/06/2022
							F08	Hospitalization or Prolonged Hospitalization	A0705	Battery Problem	13/06/2022	13/06/2022
									A070504	Premature Discharge of Battery	13/06/2022	13/06/2022
							F1905	Device Revision or Replacement	A0705	Battery Problem	13/06/2022	13/06/2022
									A070504	Premature Discharge of Battery	13/06/2022	13/06/2022
					CT981	Single use	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem	13/06/2022	13/06/2022
									A070504	Premature Discharge of Battery	13/06/2022	13/06/2022
							F08	Hospitalization or Prolonged Hospitalization	A0705	Battery Problem	13/06/2022	13/06/2022
									A070504	Premature Discharge of Battery	13/06/2022	13/06/2022
							F1905	Device Revision or Replacement	A0705	Battery Problem	13/06/2022	13/06/2022
									A070504	Premature Discharge of Battery	13/06/2022	13/06/2022
					CT983	Surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem	13/06/2022	13/06/2022
									A070504	Premature Discharge of Battery	13/06/2022	13/06/2022
							F08	Hospitalization or Prolonged Hospitalization	A0705	Battery Problem	13/06/2022	13/06/2022
									A070504	Premature Discharge of Battery	13/06/2022	13/06/2022
							F1905	Device Revision or Replacement	A0705	Battery Problem	13/06/2022	13/06/2022
									A070504	Premature Discharge of Battery	13/06/2022	13/06/2022
7883	2022	All other reportable incidents	35852	Single-chamber implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem	11/04/2022	11/04/2022
									A070504	Premature Discharge of Battery	11/04/2022	11/04/2022
							F08	Hospitalization or Prolonged Hospitalization	A0705	Battery Problem	11/04/2022	11/04/2022
									A070504	Premature Discharge of Battery	11/04/2022	11/04/2022
							F1905	Device Revision or Replacement	A0705	Battery Problem	11/04/2022	11/04/2022
									A070504	Premature Discharge of Battery	11/04/2022	11/04/2022
7884	2022	All other reportable incidents	35852	Single-chamber implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	15/03/2022	15/03/2022
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	15/03/2022	15/03/2022
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	15/03/2022	15/03/2022
7885	2022	All other reportable incidents	35852	Single-chamber implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	01/03/2022	01/03/2022
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	01/03/2022	01/03/2022
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	01/03/2022	01/03/2022
7886	2022		47265	Dual-chamber implantable pacemaker, rate-responsive	CT141	Electrodes/Leads and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	28/09/2022	28/09/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	28/09/2022	28/09/2022
					CT145	Cardiovascular devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	28/09/2022	28/09/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	28/09/2022	28/09/2022
					CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	28/09/2022	28/09/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	28/09/2022	28/09/2022
					CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	28/09/2022	28/09/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	28/09/2022	28/09/2022
					CT233	Surgical	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	28/09/2022	28/09/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	28/09/2022	28/09/2022
					CT2406	Implantable	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	28/09/2022	28/09/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	28/09/2022	28/09/2022
					CT2407	Active implantable	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	28/09/2022	28/09/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	28/09/2022	28/09/2022
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	28/09/2022	28/09/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	28/09/2022	28/09/2022
					CT2727	Active implant devices/ancillaries	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	28/09/2022	28/09/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	28/09/2022	28/09/2022
					CT2750	Implantable cardiac electrical stimulators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	28/09/2022	28/09/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	28/09/2022	28/09/2022
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	28/09/2022	28/09/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	28/09/2022	28/09/2022
					CT2985	Active	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	28/09/2022	28/09/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	28/09/2022	28/09/2022
					CT3001	Implantable metal	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	28/09/2022	28/09/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	28/09/2022	28/09/2022
					CT3002	Implantable synthetic polymer	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	28/09/2022	28/09/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	28/09/2022	28/09/2022
					CT3006	Implantable inorganic material	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	28/09/2022	28/09/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	28/09/2022	28/09/2022
					CT314	Electrically powered	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	28/09/2022	28/09/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	28/09/2022	28/09/2022
					CT321	Long-term surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	28/09/2022	28/09/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	28/09/2022	28/09/2022
					CT449	Generators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	28/09/2022	28/09/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	28/09/2022	28/09/2022
					CT981	Single use	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	28/09/2022	28/09/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	28/09/2022	28/09/2022
					CT983	Surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	28/09/2022	28/09/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	28/09/2022	28/09/2022
7887	2022	All other reportable incidents	47265	Dual-chamber implantable pacemaker, rate-responsive	CT141	Electrodes/Leads and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	24/09/2022	24/09/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	24/09/2022	24/09/2022
					CT145	Cardiovascular devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	24/09/2022	24/09/2022

							F1901	Additional Surgery	A24	No Apparent Device Problem	24/09/2022	24/09/2022
					CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	24/09/2022	24/09/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	24/09/2022	24/09/2022
					CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	24/09/2022	24/09/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	24/09/2022	24/09/2022
					CT233	Surgical	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	24/09/2022	24/09/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	24/09/2022	24/09/2022
					CT2406	Implantable	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	24/09/2022	24/09/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	24/09/2022	24/09/2022
					CT2407	Active implantable	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	24/09/2022	24/09/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	24/09/2022	24/09/2022
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	24/09/2022	24/09/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	24/09/2022	24/09/2022
					CT2727	Active implant devices/ancillaries	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	24/09/2022	24/09/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	24/09/2022	24/09/2022
					CT2750	Implantable cardiac electrical stimulators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	24/09/2022	24/09/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	24/09/2022	24/09/2022
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	24/09/2022	24/09/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	24/09/2022	24/09/2022
					CT2985	Active	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	24/09/2022	24/09/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	24/09/2022	24/09/2022
					CT3001	Implantable metal	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	24/09/2022	24/09/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	24/09/2022	24/09/2022
					CT3002	Implantable synthetic polymer	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	24/09/2022	24/09/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	24/09/2022	24/09/2022
					CT3006	Implantable inorganic material	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	24/09/2022	24/09/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	24/09/2022	24/09/2022
					CT314	Electrically powered	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	24/09/2022	24/09/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	24/09/2022	24/09/2022
					CT321	Long-term surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	24/09/2022	24/09/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	24/09/2022	24/09/2022
					CT449	Generators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	24/09/2022	24/09/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	24/09/2022	24/09/2022
					CT981	Single use	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	24/09/2022	24/09/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	24/09/2022	24/09/2022
					CT983	Surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	24/09/2022	24/09/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	24/09/2022	24/09/2022
7888	2022	All other reportable incidents	47265	Dual-chamber implantable pacemaker, rate-responsive	CT141	Electrodes/Leads and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	24/09/2022	24/09/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	24/09/2022	24/09/2022
					CT145	Cardiovascular devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	24/09/2022	24/09/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	24/09/2022	24/09/2022
					CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	24/09/2022	24/09/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	24/09/2022	24/09/2022
					CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	24/09/2022	24/09/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	24/09/2022	24/09/2022
					CT233	Surgical	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	24/09/2022	24/09/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	24/09/2022	24/09/2022
					CT2406	Implantable	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	24/09/2022	24/09/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	24/09/2022	24/09/2022
					CT2407	Active implantable	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	24/09/2022	24/09/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	24/09/2022	24/09/2022
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	24/09/2022	24/09/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	24/09/2022	24/09/2022
					CT2727	Active implant devices/ancillaries	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	24/09/2022	24/09/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	24/09/2022	24/09/2022
					CT2750	Implantable cardiac electrical stimulators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	24/09/2022	24/09/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	24/09/2022	24/09/2022
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	24/09/2022	24/09/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	24/09/2022	24/09/2022
					CT2985	Active	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	24/09/2022	24/09/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	24/09/2022	24/09/2022
					CT3001	Implantable metal	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	24/09/2022	24/09/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	24/09/2022	24/09/2022
					CT3002	Implantable synthetic polymer	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	24/09/2022	24/09/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	24/09/2022	24/09/2022
					CT3006	Implantable inorganic material	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	24/09/2022	24/09/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	24/09/2022	24/09/2022
					CT314	Electrically powered	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	24/09/2022	24/09/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	24/09/2022	24/09/2022
					CT321	Long-term surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	24/09/2022	24/09/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	24/09/2022	24/09/2022
					CT449	Generators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	24/09/2022	24/09/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	24/09/2022	24/09/2022
					CT981	Single use	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	24/09/2022	24/09/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	24/09/2022	24/09/2022
					CT983	Surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	24/09/2022	24/09/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	24/09/2022	24/09/2022
7889	2022	Death	47265	Dual-chamber implantable pacemaker, rate-responsive	CT141	Electrodes/Leads and associated devices	E0601	Arrhythmia	A070101	Failure to Capture		
							F02	Death	A070101	Failure to Capture		
					CT145	Cardiovascular devices	E0601	Arrhythmia	A070101	Failure to Capture		
							F02	Death	A070101	Failure to Capture		
					CT1684	Cardiac pacemakers and associated devices	E0601	Arrhythmia	A070101	Failure to Capture		
							F02	Death	A070101	Failure to Capture		
					CT1685	Cardiac pacemakers	E0601	Arrhythmia	A070101	Failure to Capture		
							F02	Death	A070101	Failure to Capture		
					CT233	Surgical	E0601	Arrhythmia	A070101	Failure to Capture		
							F02	Death	A070101	Failure to Capture		
					CT2406	Implantable	E0601	Arrhythmia	A070101	Failure to Capture		

							F02	Death	A070101	Failure to Capture			
					CT2407	Active implantable	E0601	Arrhythmia	A070101	Failure to Capture			
							F02	Death	A070101	Failure to Capture			
					CT257	Implantable cardiac pacemakers	E0601	Arrhythmia	A070101	Failure to Capture			
							F02	Death	A070101	Failure to Capture			
					CT2727	Active implant devices/ancillaries	E0601	Arrhythmia	A070101	Failure to Capture			
							F02	Death	A070101	Failure to Capture			
					CT2750	Implantable cardiac electrical stimulators and associated devices	E0601	Arrhythmia	A070101	Failure to Capture			
							F02	Death	A070101	Failure to Capture			
					CT2866	Implantable cardiac electrical stimulators	E0601	Arrhythmia	A070101	Failure to Capture			
							F02	Death	A070101	Failure to Capture			
					CT2985	Active	E0601	Arrhythmia	A070101	Failure to Capture			
							F02	Death	A070101	Failure to Capture			
					CT3001	Implantable metal	E0601	Arrhythmia	A070101	Failure to Capture			
							F02	Death	A070101	Failure to Capture			
					CT3002	Implantable synthetic polymer	E0601	Arrhythmia	A070101	Failure to Capture			
							F02	Death	A070101	Failure to Capture			
					CT3006	Implantable inorganic material	E0601	Arrhythmia	A070101	Failure to Capture			
							F02	Death	A070101	Failure to Capture			
					CT314	Electrically powered	E0601	Arrhythmia	A070101	Failure to Capture			
							F02	Death	A070101	Failure to Capture			
					CT321	Long-term surgical invasive	E0601	Arrhythmia	A070101	Failure to Capture			
							F02	Death	A070101	Failure to Capture			
					CT449	Generators and associated devices	E0601	Arrhythmia	A070101	Failure to Capture			
							F02	Death	A070101	Failure to Capture			
					CT981	Single use	E0601	Arrhythmia	A070101	Failure to Capture			
							F02	Death	A070101	Failure to Capture			
					CT983	Surgical invasive	E0601	Arrhythmia	A070101	Failure to Capture			
							F02	Death	A070101	Failure to Capture			
7890	2022	All other reportable incidents	47270	Cardiac resynchronization therapy implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	05/02/2022	05/02/2022	
							F1901	Additional Surgery	A070504	Premature Discharge of Battery	05/02/2022	05/02/2022	
7891	2022	All other reportable incidents	47265	Dual-chamber implantable pacemaker, rate-responsive	CT141	Electrodes/Leads and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A071206	Pacing Asynchronously			
							F26	No Health Consequences or Impact	A071206	Pacing Asynchronously			
					CT145	Cardiovascular devices	E2403	No Clinical Signs, Symptoms or Conditions	A071206	Pacing Asynchronously			
							F26	No Health Consequences or Impact	A071206	Pacing Asynchronously			
					CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A071206	Pacing Asynchronously			
							F26	No Health Consequences or Impact	A071206	Pacing Asynchronously			
					CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A071206	Pacing Asynchronously			
							F26	No Health Consequences or Impact	A071206	Pacing Asynchronously			
					CT233	Surgical	E2403	No Clinical Signs, Symptoms or Conditions	A071206	Pacing Asynchronously			
							F26	No Health Consequences or Impact	A071206	Pacing Asynchronously			
					CT2406	Implantable	E2403	No Clinical Signs, Symptoms or Conditions	A071206	Pacing Asynchronously			
							F26	No Health Consequences or Impact	A071206	Pacing Asynchronously			
					CT2407	Active implantable	E2403	No Clinical Signs, Symptoms or Conditions	A071206	Pacing Asynchronously			
							F26	No Health Consequences or Impact	A071206	Pacing Asynchronously			
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A071206	Pacing Asynchronously			
							F26	No Health Consequences or Impact	A071206	Pacing Asynchronously			
					CT2727	Active implant devices/ancillaries	E2403	No Clinical Signs, Symptoms or Conditions	A071206	Pacing Asynchronously			
							F26	No Health Consequences or Impact	A071206	Pacing Asynchronously			
					CT2750	Implantable cardiac electrical stimulators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A071206	Pacing Asynchronously			
							F26	No Health Consequences or Impact	A071206	Pacing Asynchronously			
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A071206	Pacing Asynchronously			
							F26	No Health Consequences or Impact	A071206	Pacing Asynchronously			
					CT2985	Active	E2403	No Clinical Signs, Symptoms or Conditions	A071206	Pacing Asynchronously			
							F26	No Health Consequences or Impact	A071206	Pacing Asynchronously			
					CT3001	Implantable metal	E2403	No Clinical Signs, Symptoms or Conditions	A071206	Pacing Asynchronously			
							F26	No Health Consequences or Impact	A071206	Pacing Asynchronously			
					CT3002	Implantable synthetic polymer	E2403	No Clinical Signs, Symptoms or Conditions	A071206	Pacing Asynchronously			
							F26	No Health Consequences or Impact	A071206	Pacing Asynchronously			
					CT3006	Implantable inorganic material	E2403	No Clinical Signs, Symptoms or Conditions	A071206	Pacing Asynchronously			
							F26	No Health Consequences or Impact	A071206	Pacing Asynchronously			
					CT314	Electrically powered	E2403	No Clinical Signs, Symptoms or Conditions	A071206	Pacing Asynchronously			
							F26	No Health Consequences or Impact	A071206	Pacing Asynchronously			
					CT321	Long-term surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A071206	Pacing Asynchronously			
							F26	No Health Consequences or Impact	A071206	Pacing Asynchronously			
					CT449	Generators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A071206	Pacing Asynchronously			
							F26	No Health Consequences or Impact	A071206	Pacing Asynchronously			
					CT981	Single use	E2403	No Clinical Signs, Symptoms or Conditions	A071206	Pacing Asynchronously			
							F26	No Health Consequences or Impact	A071206	Pacing Asynchronously			
					CT983	Surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A071206	Pacing Asynchronously			
							F26	No Health Consequences or Impact	A071206	Pacing Asynchronously			
7892	2022	All other reportable incidents	47265	Dual-chamber implantable pacemaker, rate-responsive	CT141	Electrodes/Leads and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070101	Failure to Capture			
									A070909	Oversensing			
							F26	No Health Consequences or Impact	A070101	Failure to Capture			
									A070909	Oversensing			
					CT145	Cardiovascular devices	E2403	No Clinical Signs, Symptoms or Conditions	A070101	Failure to Capture			
									A070909	Oversensing			
							F26	No Health Consequences or Impact	A070101	Failure to Capture			
									A070909	Oversensing			
					CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070101	Failure to Capture			
									A070909	Oversensing			
							F26	No Health Consequences or Impact	A070101	Failure to Capture			
									A070909	Oversensing			
					CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A070101	Failure to Capture			
									A070909	Oversensing			
							F26	No Health Consequences or Impact	A070101	Failure to Capture			
									A070909	Oversensing			
					CT233	Surgical	E2403	No Clinical Signs, Symptoms or Conditions	A070101	Failure to Capture			

									A070909	Oversensing		
							F26	No Health Consequences or Impact	A070101	Failure to Capture		
									A070909	Oversensing		
					CT2406	Implantable	E2403	No Clinical Signs, Symptoms or Conditions	A070101	Failure to Capture		
									A070909	Oversensing		
							F26	No Health Consequences or Impact	A070101	Failure to Capture		
									A070909	Oversensing		
					CT2407	Active implantable	E2403	No Clinical Signs, Symptoms or Conditions	A070101	Failure to Capture		
									A070909	Oversensing		
							F26	No Health Consequences or Impact	A070101	Failure to Capture		
									A070909	Oversensing		
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A070101	Failure to Capture		
									A070909	Oversensing		
							F26	No Health Consequences or Impact	A070101	Failure to Capture		
									A070909	Oversensing		
					CT2727	Active implant devices/ancillaries	E2403	No Clinical Signs, Symptoms or Conditions	A070101	Failure to Capture		
									A070909	Oversensing		
							F26	No Health Consequences or Impact	A070101	Failure to Capture		
									A070909	Oversensing		
					CT2750	Implantable cardiac electrical stimulators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070101	Failure to Capture		
									A070909	Oversensing		
							F26	No Health Consequences or Impact	A070101	Failure to Capture		
									A070909	Oversensing		
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A070101	Failure to Capture		
									A070909	Oversensing		
							F26	No Health Consequences or Impact	A070101	Failure to Capture		
									A070909	Oversensing		
					CT2985	Active	E2403	No Clinical Signs, Symptoms or Conditions	A070101	Failure to Capture		
									A070909	Oversensing		
							F26	No Health Consequences or Impact	A070101	Failure to Capture		
									A070909	Oversensing		
					CT3001	Implantable metal	E2403	No Clinical Signs, Symptoms or Conditions	A070101	Failure to Capture		
									A070909	Oversensing		
							F26	No Health Consequences or Impact	A070101	Failure to Capture		
									A070909	Oversensing		
					CT3002	Implantable synthetic polymer	E2403	No Clinical Signs, Symptoms or Conditions	A070101	Failure to Capture		
									A070909	Oversensing		
							F26	No Health Consequences or Impact	A070101	Failure to Capture		
									A070909	Oversensing		
					CT3006	Implantable inorganic material	E2403	No Clinical Signs, Symptoms or Conditions	A070101	Failure to Capture		
									A070909	Oversensing		
							F26	No Health Consequences or Impact	A070101	Failure to Capture		
									A070909	Oversensing		
					CT314	Electrically powered	E2403	No Clinical Signs, Symptoms or Conditions	A070101	Failure to Capture		
									A070909	Oversensing		
							F26	No Health Consequences or Impact	A070101	Failure to Capture		
									A070909	Oversensing		
					CT321	Long-term surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A070101	Failure to Capture		
									A070909	Oversensing		
							F26	No Health Consequences or Impact	A070101	Failure to Capture		
									A070909	Oversensing		
					CT449	Generators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070101	Failure to Capture		
									A070909	Oversensing		
							F26	No Health Consequences or Impact	A070101	Failure to Capture		
									A070909	Oversensing		
					CT981	Single use	E2403	No Clinical Signs, Symptoms or Conditions	A070101	Failure to Capture		
									A070909	Oversensing		
							F26	No Health Consequences or Impact	A070101	Failure to Capture		
									A070909	Oversensing		
					CT983	Surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A070101	Failure to Capture		
									A070909	Oversensing		
							F26	No Health Consequences or Impact	A070101	Failure to Capture		
									A070909	Oversensing		
7893	2022	All other reportable incidents	35852	Single-chamber implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E0601	Arrhythmia	A070909	Oversensing		
									A071302	Inappropriate/Inadequate Shock/Stimulation		
							E2311	Discomfort	A070909	Oversensing		
									A071302	Inappropriate/Inadequate Shock/Stimulation		
							E2403	No Clinical Signs, Symptoms or Conditions	A070909	Oversensing		
									A071302	Inappropriate/Inadequate Shock/Stimulation		
							F23	Unexpected Medical Intervention	A070909	Oversensing		
									A071302	Inappropriate/Inadequate Shock/Stimulation		
7894	2022	All other reportable incidents	17882	Physiologic-monitoring defibrillation system	CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070803	Failure to Power Up		
							F26	No Health Consequences or Impact	A070803	Failure to Power Up		
7895	2022	All other reportable incidents	47267	Single-chamber implantable pacemaker, rate-responsive	CT141	Electrodes/Leads and associated devices	E0112	Dizziness	A070504	Premature Discharge of Battery	13/10/2022	13/10/2022
							E060102	Atrial Fibrillation	A070504	Premature Discharge of Battery	13/10/2022	13/10/2022
							E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	13/10/2022	13/10/2022
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	13/10/2022	13/10/2022
							F12	Serious Injury/ Illness/ Impairment	A070504	Premature Discharge of Battery	13/10/2022	13/10/2022
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	13/10/2022	13/10/2022
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery	13/10/2022	13/10/2022
					CT145	Cardiovascular devices	E0112	Dizziness	A070504	Premature Discharge of Battery	13/10/2022	13/10/2022
							E060102	Atrial Fibrillation	A070504	Premature Discharge of Battery	13/10/2022	13/10/2022
							E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	13/10/2022	13/10/2022
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	13/10/2022	13/10/2022
							F12	Serious Injury/ Illness/ Impairment	A070504	Premature Discharge of Battery	13/10/2022	13/10/2022
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	13/10/2022	13/10/2022
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery	13/10/2022	13/10/2022
					CT1684	Cardiac pacemakers and associated devices	E0112	Dizziness	A070504	Premature Discharge of Battery	13/10/2022	13/10/2022

							E060102	Atrial Fibrillation	A070504	Premature Discharge of Battery	13/10/2022	13/10/2022
							E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	13/10/2022	13/10/2022
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	13/10/2022	13/10/2022
							F12	Serious Injury/ Illness/ Impairment	A070504	Premature Discharge of Battery	13/10/2022	13/10/2022
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	13/10/2022	13/10/2022
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery	13/10/2022	13/10/2022
					CT1685	Cardiac pacemakers	E0112	Dizziness	A070504	Premature Discharge of Battery	13/10/2022	13/10/2022
							E060102	Atrial Fibrillation	A070504	Premature Discharge of Battery	13/10/2022	13/10/2022
							E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	13/10/2022	13/10/2022
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	13/10/2022	13/10/2022
							F12	Serious Injury/ Illness/ Impairment	A070504	Premature Discharge of Battery	13/10/2022	13/10/2022
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	13/10/2022	13/10/2022
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery	13/10/2022	13/10/2022
					CT233	Surgical	E0112	Dizziness	A070504	Premature Discharge of Battery	13/10/2022	13/10/2022
							E060102	Atrial Fibrillation	A070504	Premature Discharge of Battery	13/10/2022	13/10/2022
							E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	13/10/2022	13/10/2022
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	13/10/2022	13/10/2022
							F12	Serious Injury/ Illness/ Impairment	A070504	Premature Discharge of Battery	13/10/2022	13/10/2022
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	13/10/2022	13/10/2022
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery	13/10/2022	13/10/2022
					CT2406	Implantable	E0112	Dizziness	A070504	Premature Discharge of Battery	13/10/2022	13/10/2022
							E060102	Atrial Fibrillation	A070504	Premature Discharge of Battery	13/10/2022	13/10/2022
							E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	13/10/2022	13/10/2022
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	13/10/2022	13/10/2022
							F12	Serious Injury/ Illness/ Impairment	A070504	Premature Discharge of Battery	13/10/2022	13/10/2022
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	13/10/2022	13/10/2022
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery	13/10/2022	13/10/2022
					CT2407	Active implantable	E0112	Dizziness	A070504	Premature Discharge of Battery	13/10/2022	13/10/2022
							E060102	Atrial Fibrillation	A070504	Premature Discharge of Battery	13/10/2022	13/10/2022
							E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	13/10/2022	13/10/2022
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	13/10/2022	13/10/2022
							F12	Serious Injury/ Illness/ Impairment	A070504	Premature Discharge of Battery	13/10/2022	13/10/2022
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	13/10/2022	13/10/2022
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery	13/10/2022	13/10/2022
					CT257	Implantable cardiac pacemakers	E0112	Dizziness	A070504	Premature Discharge of Battery	13/10/2022	13/10/2022
							E060102	Atrial Fibrillation	A070504	Premature Discharge of Battery	13/10/2022	13/10/2022
							E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	13/10/2022	13/10/2022
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	13/10/2022	13/10/2022
							F12	Serious Injury/ Illness/ Impairment	A070504	Premature Discharge of Battery	13/10/2022	13/10/2022
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	13/10/2022	13/10/2022
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery	13/10/2022	13/10/2022
					CT2727	Active implant devices/ancillaries	E0112	Dizziness	A070504	Premature Discharge of Battery	13/10/2022	13/10/2022
							E060102	Atrial Fibrillation	A070504	Premature Discharge of Battery	13/10/2022	13/10/2022
							E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	13/10/2022	13/10/2022
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	13/10/2022	13/10/2022
							F12	Serious Injury/ Illness/ Impairment	A070504	Premature Discharge of Battery	13/10/2022	13/10/2022
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	13/10/2022	13/10/2022
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery	13/10/2022	13/10/2022
					CT2750	Implantable cardiac electrical stimulators and associated devices	E0112	Dizziness	A070504	Premature Discharge of Battery	13/10/2022	13/10/2022
							E060102	Atrial Fibrillation	A070504	Premature Discharge of Battery	13/10/2022	13/10/2022
							E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	13/10/2022	13/10/2022
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	13/10/2022	13/10/2022
							F12	Serious Injury/ Illness/ Impairment	A070504	Premature Discharge of Battery	13/10/2022	13/10/2022
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	13/10/2022	13/10/2022
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery	13/10/2022	13/10/2022
					CT2866	Implantable cardiac electrical stimulators	E0112	Dizziness	A070504	Premature Discharge of Battery	13/10/2022	13/10/2022
							E060102	Atrial Fibrillation	A070504	Premature Discharge of Battery	13/10/2022	13/10/2022
							E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	13/10/2022	13/10/2022
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	13/10/2022	13/10/2022
							F12	Serious Injury/ Illness/ Impairment	A070504	Premature Discharge of Battery	13/10/2022	13/10/2022
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	13/10/2022	13/10/2022
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery	13/10/2022	13/10/2022
					CT2985	Active	E0112	Dizziness	A070504	Premature Discharge of Battery	13/10/2022	13/10/2022
							E060102	Atrial Fibrillation	A070504	Premature Discharge of Battery	13/10/2022	13/10/2022
							E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	13/10/2022	13/10/2022
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	13/10/2022	13/10/2022
							F12	Serious Injury/ Illness/ Impairment	A070504	Premature Discharge of Battery	13/10/2022	13/10/2022
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	13/10/2022	13/10/2022
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery	13/10/2022	13/10/2022
					CT3001	Implantable metal	E0112	Dizziness	A070504	Premature Discharge of Battery	13/10/2022	13/10/2022
							E060102	Atrial Fibrillation	A070504	Premature Discharge of Battery	13/10/2022	13/10/2022
							E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	13/10/2022	13/10/2022
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	13/10/2022	13/10/2022
							F12	Serious Injury/ Illness/ Impairment	A070504	Premature Discharge of Battery	13/10/2022	13/10/2022
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	13/10/2022	13/10/2022
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery	13/10/2022	13/10/2022
					CT3002	Implantable synthetic polymer	E0112	Dizziness	A070504	Premature Discharge of Battery	13/10/2022	13/10/2022
							E060102	Atrial Fibrillation	A070504	Premature Discharge of Battery	13/10/2022	13/10/2022
							E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	13/10/2022	13/10/2022
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	13/10/2022	13/10/2022
							F12	Serious Injury/ Illness/ Impairment	A070504	Premature Discharge of Battery	13/10/2022	13/10/2022
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	13/10/2022	13/10/2022
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery	13/10/2022	13/10/2022
					CT3006	Implantable inorganic material	E0112	Dizziness	A070504	Premature Discharge of Battery	13/10/2022	13/10/2022
							E060102	Atrial Fibrillation	A070504	Premature Discharge of Battery	13/10/2022	13/10/2022
							E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	13/10/2022	13/10/2022
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	13/10/2022	13/10/2022
							F12	Serious Injury/ Illness/ Impairment	A070504	Premature Discharge of Battery	13/10/2022	13/10/2022

							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	13/10/2022	13/10/2022
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery	13/10/2022	13/10/2022
					CT314	Electrically powered	E0112	Dizziness	A070504	Premature Discharge of Battery	13/10/2022	13/10/2022
							E060102	Atrial Fibrillation	A070504	Premature Discharge of Battery	13/10/2022	13/10/2022
							E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	13/10/2022	13/10/2022
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	13/10/2022	13/10/2022
							F12	Serious Injury/ Illness/ Impairment	A070504	Premature Discharge of Battery	13/10/2022	13/10/2022
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	13/10/2022	13/10/2022
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery	13/10/2022	13/10/2022
					CT321	Long-term surgical invasive	E0112	Dizziness	A070504	Premature Discharge of Battery	13/10/2022	13/10/2022
							E060102	Atrial Fibrillation	A070504	Premature Discharge of Battery	13/10/2022	13/10/2022
							E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	13/10/2022	13/10/2022
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	13/10/2022	13/10/2022
							F12	Serious Injury/ Illness/ Impairment	A070504	Premature Discharge of Battery	13/10/2022	13/10/2022
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	13/10/2022	13/10/2022
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery	13/10/2022	13/10/2022
					CT449	Generators and associated devices	E0112	Dizziness	A070504	Premature Discharge of Battery	13/10/2022	13/10/2022
							E060102	Atrial Fibrillation	A070504	Premature Discharge of Battery	13/10/2022	13/10/2022
							E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	13/10/2022	13/10/2022
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	13/10/2022	13/10/2022
							F12	Serious Injury/ Illness/ Impairment	A070504	Premature Discharge of Battery	13/10/2022	13/10/2022
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	13/10/2022	13/10/2022
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery	13/10/2022	13/10/2022
					CT981	Single use	E0112	Dizziness	A070504	Premature Discharge of Battery	13/10/2022	13/10/2022
							E060102	Atrial Fibrillation	A070504	Premature Discharge of Battery	13/10/2022	13/10/2022
							E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	13/10/2022	13/10/2022
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	13/10/2022	13/10/2022
							F12	Serious Injury/ Illness/ Impairment	A070504	Premature Discharge of Battery	13/10/2022	13/10/2022
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	13/10/2022	13/10/2022
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery	13/10/2022	13/10/2022
					CT983	Surgical invasive	E0112	Dizziness	A070504	Premature Discharge of Battery	13/10/2022	13/10/2022
							E060102	Atrial Fibrillation	A070504	Premature Discharge of Battery	13/10/2022	13/10/2022
							E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	13/10/2022	13/10/2022
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	13/10/2022	13/10/2022
							F12	Serious Injury/ Illness/ Impairment	A070504	Premature Discharge of Battery	13/10/2022	13/10/2022
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	13/10/2022	13/10/2022
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery	13/10/2022	13/10/2022
7896	2022	All other reportable incidents	47270	Cardiac resynchronization therapy implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A110303	Program or Algorithm Execution Failure		
							F26	No Health Consequences or Impact	A110303	Program or Algorithm Execution Failure		
7897	2022	All other reportable incidents	35852	Single-chamber implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem	13/12/2022	13/12/2022
									A070504	Premature Discharge of Battery	13/12/2022	13/12/2022
									A0908	Incorrect, Inadequate or Imprecise Result or Readings	13/12/2022	13/12/2022
							F08	Hospitalization or Prolonged Hospitalization	A0705	Battery Problem	13/12/2022	13/12/2022
									A070504	Premature Discharge of Battery	13/12/2022	13/12/2022
									A0908	Incorrect, Inadequate or Imprecise Result or Readings	13/12/2022	13/12/2022
							F1905	Device Revision or Replacement	A0705	Battery Problem	13/12/2022	13/12/2022
									A070504	Premature Discharge of Battery	13/12/2022	13/12/2022
									A0908	Incorrect, Inadequate or Imprecise Result or Readings	13/12/2022	13/12/2022
7898	2022	All other reportable incidents	37265	Dual-chamber implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A0712	Pacing Problem		
							F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A0712	Pacing Problem		
7899	2022	All other reportable incidents	47265	Dual-chamber implantable pacemaker, rate-responsive	CT141	Electrodes/Leads and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	24/09/2022	24/09/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	24/09/2022	24/09/2022
					CT145	Cardiovascular devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	24/09/2022	24/09/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	24/09/2022	24/09/2022
					CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	24/09/2022	24/09/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	24/09/2022	24/09/2022
					CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	24/09/2022	24/09/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	24/09/2022	24/09/2022
					CT233	Surgical	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	24/09/2022	24/09/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	24/09/2022	24/09/2022
					CT2406	Implantable	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	24/09/2022	24/09/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	24/09/2022	24/09/2022
					CT2407	Active implantable	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	24/09/2022	24/09/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	24/09/2022	24/09/2022
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	24/09/2022	24/09/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	24/09/2022	24/09/2022
					CT2727	Active implant devices/ancillaries	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	24/09/2022	24/09/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	24/09/2022	24/09/2022
					CT2750	Implantable cardiac electrical stimulators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	24/09/2022	24/09/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	24/09/2022	24/09/2022
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	24/09/2022	24/09/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	24/09/2022	24/09/2022
					CT2985	Active	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	24/09/2022	24/09/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	24/09/2022	24/09/2022
					CT3001	Implantable metal	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	24/09/2022	24/09/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	24/09/2022	24/09/2022
					CT3002	Implantable synthetic polymer	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	24/09/2022	24/09/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	24/09/2022	24/09/2022
					CT3006	Implantable inorganic material	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	24/09/2022	24/09/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	24/09/2022	24/09/2022
					CT314	Electrically powered	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	24/09/2022	24/09/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	24/09/2022	24/09/2022
					CT321	Long-term surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	24/09/2022	24/09/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	24/09/2022	24/09/2022
					CT449	Generators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	24/09/2022	24/09/2022

						F1901	Additional Surgery	A24	No Apparent Device Problem	24/09/2022	24/09/2022
					CT981	Single use	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	24/09/2022	24/09/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	24/09/2022	24/09/2022
					CT983	Surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	24/09/2022	24/09/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	24/09/2022	24/09/2022
7900	2022	All other reportable incidents	47265	Dual-chamber implantable pacemaker, rate-responsive	CT141	Electrodes/Leads and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem	01/03/2022	01/03/2022
									A070504	Premature Discharge of Battery	01/03/2022	01/03/2022
							F08	Hospitalization or Prolonged Hospitalization	A0705	Battery Problem	01/03/2022	01/03/2022
									A070504	Premature Discharge of Battery	01/03/2022	01/03/2022
							F1905	Device Revision or Replacement	A0705	Battery Problem	01/03/2022	01/03/2022
									A070504	Premature Discharge of Battery	01/03/2022	01/03/2022
					CT145	Cardiovascular devices	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem	01/03/2022	01/03/2022
									A070504	Premature Discharge of Battery	01/03/2022	01/03/2022
							F08	Hospitalization or Prolonged Hospitalization	A0705	Battery Problem	01/03/2022	01/03/2022
									A070504	Premature Discharge of Battery	01/03/2022	01/03/2022
							F1905	Device Revision or Replacement	A0705	Battery Problem	01/03/2022	01/03/2022
									A070504	Premature Discharge of Battery	01/03/2022	01/03/2022
					CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem	01/03/2022	01/03/2022
									A070504	Premature Discharge of Battery	01/03/2022	01/03/2022
							F08	Hospitalization or Prolonged Hospitalization	A0705	Battery Problem	01/03/2022	01/03/2022
									A070504	Premature Discharge of Battery	01/03/2022	01/03/2022
							F1905	Device Revision or Replacement	A0705	Battery Problem	01/03/2022	01/03/2022
									A070504	Premature Discharge of Battery	01/03/2022	01/03/2022
					CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem	01/03/2022	01/03/2022
									A070504	Premature Discharge of Battery	01/03/2022	01/03/2022
							F08	Hospitalization or Prolonged Hospitalization	A0705	Battery Problem	01/03/2022	01/03/2022
									A070504	Premature Discharge of Battery	01/03/2022	01/03/2022
							F1905	Device Revision or Replacement	A0705	Battery Problem	01/03/2022	01/03/2022
									A070504	Premature Discharge of Battery	01/03/2022	01/03/2022
					CT233	Surgical	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem	01/03/2022	01/03/2022
									A070504	Premature Discharge of Battery	01/03/2022	01/03/2022
							F08	Hospitalization or Prolonged Hospitalization	A0705	Battery Problem	01/03/2022	01/03/2022
									A070504	Premature Discharge of Battery	01/03/2022	01/03/2022
							F1905	Device Revision or Replacement	A0705	Battery Problem	01/03/2022	01/03/2022
									A070504	Premature Discharge of Battery	01/03/2022	01/03/2022
					CT2406	Implantable	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem	01/03/2022	01/03/2022
									A070504	Premature Discharge of Battery	01/03/2022	01/03/2022
							F08	Hospitalization or Prolonged Hospitalization	A0705	Battery Problem	01/03/2022	01/03/2022
									A070504	Premature Discharge of Battery	01/03/2022	01/03/2022
							F1905	Device Revision or Replacement	A0705	Battery Problem	01/03/2022	01/03/2022
									A070504	Premature Discharge of Battery	01/03/2022	01/03/2022
					CT2407	Active implantable	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem	01/03/2022	01/03/2022
									A070504	Premature Discharge of Battery	01/03/2022	01/03/2022
							F08	Hospitalization or Prolonged Hospitalization	A0705	Battery Problem	01/03/2022	01/03/2022
									A070504	Premature Discharge of Battery	01/03/2022	01/03/2022
							F1905	Device Revision or Replacement	A0705	Battery Problem	01/03/2022	01/03/2022
									A070504	Premature Discharge of Battery	01/03/2022	01/03/2022
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem	01/03/2022	01/03/2022
									A070504	Premature Discharge of Battery	01/03/2022	01/03/2022
							F08	Hospitalization or Prolonged Hospitalization	A0705	Battery Problem	01/03/2022	01/03/2022
									A070504	Premature Discharge of Battery	01/03/2022	01/03/2022
							F1905	Device Revision or Replacement	A0705	Battery Problem	01/03/2022	01/03/2022
									A070504	Premature Discharge of Battery	01/03/2022	01/03/2022
					CT2727	Active implant devices/ancillaries	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem	01/03/2022	01/03/2022
									A070504	Premature Discharge of Battery	01/03/2022	01/03/2022
							F08	Hospitalization or Prolonged Hospitalization	A0705	Battery Problem	01/03/2022	01/03/2022
									A070504	Premature Discharge of Battery	01/03/2022	01/03/2022
							F1905	Device Revision or Replacement	A0705	Battery Problem	01/03/2022	01/03/2022
									A070504	Premature Discharge of Battery	01/03/2022	01/03/2022
					CT2750	Implantable cardiac electrical stimulators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem	01/03/2022	01/03/2022
									A070504	Premature Discharge of Battery	01/03/2022	01/03/2022
							F08	Hospitalization or Prolonged Hospitalization	A0705	Battery Problem	01/03/2022	01/03/2022
									A070504	Premature Discharge of Battery	01/03/2022	01/03/2022
							F1905	Device Revision or Replacement	A0705	Battery Problem	01/03/2022	01/03/2022
									A070504	Premature Discharge of Battery	01/03/2022	01/03/2022
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem	01/03/2022	01/03/2022
									A070504	Premature Discharge of Battery	01/03/2022	01/03/2022
							F08	Hospitalization or Prolonged Hospitalization	A0705	Battery Problem	01/03/2022	01/03/2022
									A070504	Premature Discharge of Battery	01/03/2022	01/03/2022
							F1905	Device Revision or Replacement	A0705	Battery Problem	01/03/2022	01/03/2022
									A070504	Premature Discharge of Battery	01/03/2022	01/03/2022
					CT2985	Active	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem	01/03/2022	01/03/2022
									A070504	Premature Discharge of Battery	01/03/2022	01/03/2022
							F08	Hospitalization or Prolonged Hospitalization	A0705	Battery Problem	01/03/2022	01/03/2022
									A070504	Premature Discharge of Battery	01/03/2022	01/03/2022
							F1905	Device Revision or Replacement	A0705	Battery Problem	01/03/2022	01/03/2022
									A070504	Premature Discharge of Battery	01/03/2022	01/03/2022
					CT3001	Implantable metal	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem	01/03/2022	01/03/2022
									A070504	Premature Discharge of Battery	01/03/2022	01/03/2022
							F08	Hospitalization or Prolonged Hospitalization	A0705	Battery Problem	01/03/2022	01/03/2022
									A070504	Premature Discharge of Battery	01/03/2022	01/03/2022
							F1905	Device Revision or Replacement	A0705	Battery Problem	01/03/2022	01/03/2022
									A070504	Premature Discharge of Battery	01/03/2022	01/03/2022
					CT3002	Implantable synthetic polymer	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem	01/03/2022	01/03/2022
									A070504	Premature Discharge of Battery	01/03/2022	01/03/2022
							F08	Hospitalization or Prolonged Hospitalization	A0705	Battery Problem	01/03/2022	01/03/2022
									A070504	Premature Discharge of Battery	01/03/2022	01/03/2022
							F1905	Device Revision or Replacement	A0705	Battery Problem	01/03/2022	01/03/2022

									A070504	Premature Discharge of Battery	01/03/2022	01/03/2022
					CT3006	Implantable inorganic material	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem	01/03/2022	01/03/2022
									A070504	Premature Discharge of Battery	01/03/2022	01/03/2022
							F08	Hospitalization or Prolonged Hospitalization	A0705	Battery Problem	01/03/2022	01/03/2022
									A070504	Premature Discharge of Battery	01/03/2022	01/03/2022
							F1905	Device Revision or Replacement	A0705	Battery Problem	01/03/2022	01/03/2022
									A070504	Premature Discharge of Battery	01/03/2022	01/03/2022
					CT314	Electrically powered	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem	01/03/2022	01/03/2022
									A070504	Premature Discharge of Battery	01/03/2022	01/03/2022
							F08	Hospitalization or Prolonged Hospitalization	A0705	Battery Problem	01/03/2022	01/03/2022
									A070504	Premature Discharge of Battery	01/03/2022	01/03/2022
							F1905	Device Revision or Replacement	A0705	Battery Problem	01/03/2022	01/03/2022
									A070504	Premature Discharge of Battery	01/03/2022	01/03/2022
					CT321	Long-term surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem	01/03/2022	01/03/2022
									A070504	Premature Discharge of Battery	01/03/2022	01/03/2022
							F08	Hospitalization or Prolonged Hospitalization	A0705	Battery Problem	01/03/2022	01/03/2022
									A070504	Premature Discharge of Battery	01/03/2022	01/03/2022
							F1905	Device Revision or Replacement	A0705	Battery Problem	01/03/2022	01/03/2022
									A070504	Premature Discharge of Battery	01/03/2022	01/03/2022
					CT449	Generators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem	01/03/2022	01/03/2022
									A070504	Premature Discharge of Battery	01/03/2022	01/03/2022
							F08	Hospitalization or Prolonged Hospitalization	A0705	Battery Problem	01/03/2022	01/03/2022
									A070504	Premature Discharge of Battery	01/03/2022	01/03/2022
							F1905	Device Revision or Replacement	A0705	Battery Problem	01/03/2022	01/03/2022
									A070504	Premature Discharge of Battery	01/03/2022	01/03/2022
					CT981	Single use	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem	01/03/2022	01/03/2022
									A070504	Premature Discharge of Battery	01/03/2022	01/03/2022
							F08	Hospitalization or Prolonged Hospitalization	A0705	Battery Problem	01/03/2022	01/03/2022
									A070504	Premature Discharge of Battery	01/03/2022	01/03/2022
							F1905	Device Revision or Replacement	A0705	Battery Problem	01/03/2022	01/03/2022
									A070504	Premature Discharge of Battery	01/03/2022	01/03/2022
					CT983	Surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem	01/03/2022	01/03/2022
									A070504	Premature Discharge of Battery	01/03/2022	01/03/2022
							F08	Hospitalization or Prolonged Hospitalization	A0705	Battery Problem	01/03/2022	01/03/2022
									A070504	Premature Discharge of Battery	01/03/2022	01/03/2022
							F1905	Device Revision or Replacement	A0705	Battery Problem	01/03/2022	01/03/2022
									A070504	Premature Discharge of Battery	01/03/2022	01/03/2022
7901	2022	All other reportable incidents	35852	Single-chamber implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A072201	High Impedance		
							F26	No Health Consequences or Impact	A072201	High Impedance		
7902	2022	All other reportable incidents	35852	Single-chamber implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem	26/10/2022	26/10/2022
									A070504	Premature Discharge of Battery	26/10/2022	26/10/2022
							F08	Hospitalization or Prolonged Hospitalization	A0705	Battery Problem	26/10/2022	26/10/2022
									A070504	Premature Discharge of Battery	26/10/2022	26/10/2022
							F1903	Device Explantation	A0705	Battery Problem	26/10/2022	26/10/2022
									A070504	Premature Discharge of Battery	26/10/2022	26/10/2022
7903	2022	All other reportable incidents	36241	Pacemaker/defibrillator lead	CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A072201	High Impedance		
							F26	No Health Consequences or Impact	A072201	High Impedance		
7904	2022	All other reportable incidents	47265	Dual-chamber implantable pacemaker, rate-responsive	CT141	Electrodes/Leads and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem	11/05/2022	11/05/2022
									A070504	Premature Discharge of Battery	11/05/2022	11/05/2022
							F08	Hospitalization or Prolonged Hospitalization	A0705	Battery Problem	11/05/2022	11/05/2022
									A070504	Premature Discharge of Battery	11/05/2022	11/05/2022
							F1903	Device Explantation	A0705	Battery Problem	11/05/2022	11/05/2022
									A070504	Premature Discharge of Battery	11/05/2022	11/05/2022
					CT145	Cardiovascular devices	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem	11/05/2022	11/05/2022
									A070504	Premature Discharge of Battery	11/05/2022	11/05/2022
							F08	Hospitalization or Prolonged Hospitalization	A0705	Battery Problem	11/05/2022	11/05/2022
									A070504	Premature Discharge of Battery	11/05/2022	11/05/2022
							F1903	Device Explantation	A0705	Battery Problem	11/05/2022	11/05/2022
									A070504	Premature Discharge of Battery	11/05/2022	11/05/2022
					CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem	11/05/2022	11/05/2022
									A070504	Premature Discharge of Battery	11/05/2022	11/05/2022
							F08	Hospitalization or Prolonged Hospitalization	A0705	Battery Problem	11/05/2022	11/05/2022
									A070504	Premature Discharge of Battery	11/05/2022	11/05/2022
							F1903	Device Explantation	A0705	Battery Problem	11/05/2022	11/05/2022
									A070504	Premature Discharge of Battery	11/05/2022	11/05/2022
					CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem	11/05/2022	11/05/2022
									A070504	Premature Discharge of Battery	11/05/2022	11/05/2022
							F08	Hospitalization or Prolonged Hospitalization	A0705	Battery Problem	11/05/2022	11/05/2022
									A070504	Premature Discharge of Battery	11/05/2022	11/05/2022
							F1903	Device Explantation	A0705	Battery Problem	11/05/2022	11/05/2022
									A070504	Premature Discharge of Battery	11/05/2022	11/05/2022
					CT233	Surgical	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem	11/05/2022	11/05/2022
									A070504	Premature Discharge of Battery	11/05/2022	11/05/2022
							F08	Hospitalization or Prolonged Hospitalization	A0705	Battery Problem	11/05/2022	11/05/2022
									A070504	Premature Discharge of Battery	11/05/2022	11/05/2022
							F1903	Device Explantation	A0705	Battery Problem	11/05/2022	11/05/2022
									A070504	Premature Discharge of Battery	11/05/2022	11/05/2022
					CT2406	Implantable	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem	11/05/2022	11/05/2022
									A070504	Premature Discharge of Battery	11/05/2022	11/05/2022
							F08	Hospitalization or Prolonged Hospitalization	A0705	Battery Problem	11/05/2022	11/05/2022
									A070504	Premature Discharge of Battery	11/05/2022	11/05/2022
							F1903	Device Explantation	A0705	Battery Problem	11/05/2022	11/05/2022
									A070504	Premature Discharge of Battery	11/05/2022	11/05/2022
					CT2407	Active implantable	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem	11/05/2022	11/05/2022
									A070504	Premature Discharge of Battery	11/05/2022	11/05/2022
							F08	Hospitalization or Prolonged Hospitalization	A0705	Battery Problem	11/05/2022	11/05/2022
									A070504	Premature Discharge of Battery	11/05/2022	11/05/2022
							F1903	Device Explantation	A0705	Battery Problem	11/05/2022	11/05/2022

									A070504	Premature Discharge of Battery	11/05/2022	11/05/2022
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem	11/05/2022	11/05/2022
									A070504	Premature Discharge of Battery	11/05/2022	11/05/2022
							F08	Hospitalization or Prolonged Hospitalization	A0705	Battery Problem	11/05/2022	11/05/2022
									A070504	Premature Discharge of Battery	11/05/2022	11/05/2022
							F1903	Device Explantation	A0705	Battery Problem	11/05/2022	11/05/2022
									A070504	Premature Discharge of Battery	11/05/2022	11/05/2022
					CT2727	Active implant devices/ancillaries	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem	11/05/2022	11/05/2022
									A070504	Premature Discharge of Battery	11/05/2022	11/05/2022
							F08	Hospitalization or Prolonged Hospitalization	A0705	Battery Problem	11/05/2022	11/05/2022
									A070504	Premature Discharge of Battery	11/05/2022	11/05/2022
							F1903	Device Explantation	A0705	Battery Problem	11/05/2022	11/05/2022
									A070504	Premature Discharge of Battery	11/05/2022	11/05/2022
					CT2750	Implantable cardiac electrical stimulators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem	11/05/2022	11/05/2022
									A070504	Premature Discharge of Battery	11/05/2022	11/05/2022
							F08	Hospitalization or Prolonged Hospitalization	A0705	Battery Problem	11/05/2022	11/05/2022
									A070504	Premature Discharge of Battery	11/05/2022	11/05/2022
							F1903	Device Explantation	A0705	Battery Problem	11/05/2022	11/05/2022
									A070504	Premature Discharge of Battery	11/05/2022	11/05/2022
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem	11/05/2022	11/05/2022
									A070504	Premature Discharge of Battery	11/05/2022	11/05/2022
							F08	Hospitalization or Prolonged Hospitalization	A0705	Battery Problem	11/05/2022	11/05/2022
									A070504	Premature Discharge of Battery	11/05/2022	11/05/2022
							F1903	Device Explantation	A0705	Battery Problem	11/05/2022	11/05/2022
									A070504	Premature Discharge of Battery	11/05/2022	11/05/2022
					CT2985	Active	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem	11/05/2022	11/05/2022
									A070504	Premature Discharge of Battery	11/05/2022	11/05/2022
							F08	Hospitalization or Prolonged Hospitalization	A0705	Battery Problem	11/05/2022	11/05/2022
									A070504	Premature Discharge of Battery	11/05/2022	11/05/2022
							F1903	Device Explantation	A0705	Battery Problem	11/05/2022	11/05/2022
									A070504	Premature Discharge of Battery	11/05/2022	11/05/2022
					CT3001	Implantable metal	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem	11/05/2022	11/05/2022
									A070504	Premature Discharge of Battery	11/05/2022	11/05/2022
							F08	Hospitalization or Prolonged Hospitalization	A0705	Battery Problem	11/05/2022	11/05/2022
									A070504	Premature Discharge of Battery	11/05/2022	11/05/2022
							F1903	Device Explantation	A0705	Battery Problem	11/05/2022	11/05/2022
									A070504	Premature Discharge of Battery	11/05/2022	11/05/2022
					CT3002	Implantable synthetic polymer	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem	11/05/2022	11/05/2022
									A070504	Premature Discharge of Battery	11/05/2022	11/05/2022
							F08	Hospitalization or Prolonged Hospitalization	A0705	Battery Problem	11/05/2022	11/05/2022
									A070504	Premature Discharge of Battery	11/05/2022	11/05/2022
							F1903	Device Explantation	A0705	Battery Problem	11/05/2022	11/05/2022
									A070504	Premature Discharge of Battery	11/05/2022	11/05/2022
					CT3006	Implantable inorganic material	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem	11/05/2022	11/05/2022
									A070504	Premature Discharge of Battery	11/05/2022	11/05/2022
							F08	Hospitalization or Prolonged Hospitalization	A0705	Battery Problem	11/05/2022	11/05/2022
									A070504	Premature Discharge of Battery	11/05/2022	11/05/2022
							F1903	Device Explantation	A0705	Battery Problem	11/05/2022	11/05/2022
									A070504	Premature Discharge of Battery	11/05/2022	11/05/2022
					CT314	Electrically powered	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem	11/05/2022	11/05/2022
									A070504	Premature Discharge of Battery	11/05/2022	11/05/2022
							F08	Hospitalization or Prolonged Hospitalization	A0705	Battery Problem	11/05/2022	11/05/2022
									A070504	Premature Discharge of Battery	11/05/2022	11/05/2022
							F1903	Device Explantation	A0705	Battery Problem	11/05/2022	11/05/2022
									A070504	Premature Discharge of Battery	11/05/2022	11/05/2022
					CT321	Long-term surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem	11/05/2022	11/05/2022
									A070504	Premature Discharge of Battery	11/05/2022	11/05/2022
							F08	Hospitalization or Prolonged Hospitalization	A0705	Battery Problem	11/05/2022	11/05/2022
									A070504	Premature Discharge of Battery	11/05/2022	11/05/2022
							F1903	Device Explantation	A0705	Battery Problem	11/05/2022	11/05/2022
									A070504	Premature Discharge of Battery	11/05/2022	11/05/2022
					CT449	Generators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem	11/05/2022	11/05/2022
									A070504	Premature Discharge of Battery	11/05/2022	11/05/2022
							F08	Hospitalization or Prolonged Hospitalization	A0705	Battery Problem	11/05/2022	11/05/2022
									A070504	Premature Discharge of Battery	11/05/2022	11/05/2022
							F1903	Device Explantation	A0705	Battery Problem	11/05/2022	11/05/2022
									A070504	Premature Discharge of Battery	11/05/2022	11/05/2022
					CT981	Single use	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem	11/05/2022	11/05/2022
									A070504	Premature Discharge of Battery	11/05/2022	11/05/2022
							F08	Hospitalization or Prolonged Hospitalization	A0705	Battery Problem	11/05/2022	11/05/2022
									A070504	Premature Discharge of Battery	11/05/2022	11/05/2022
							F1903	Device Explantation	A0705	Battery Problem	11/05/2022	11/05/2022
									A070504	Premature Discharge of Battery	11/05/2022	11/05/2022
					CT983	Surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem	11/05/2022	11/05/2022
									A070504	Premature Discharge of Battery	11/05/2022	11/05/2022
							F08	Hospitalization or Prolonged Hospitalization	A0705	Battery Problem	11/05/2022	11/05/2022
									A070504	Premature Discharge of Battery	11/05/2022	11/05/2022
							F1903	Device Explantation	A0705	Battery Problem	11/05/2022	11/05/2022
									A070504	Premature Discharge of Battery	11/05/2022	11/05/2022
7905	2022	All other reportable incidents	37265	Dual-chamber implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A071102	Failure to Interrogate	26/09/2022	26/09/2022
							F08	Hospitalization or Prolonged Hospitalization	A071102	Failure to Interrogate	26/09/2022	26/09/2022
							F12	Serious Injury/ Illness/ Impairment	A071102	Failure to Interrogate	26/09/2022	26/09/2022
							F1905	Device Revision or Replacement	A071102	Failure to Interrogate	26/09/2022	26/09/2022
7906	2022	All other reportable incidents	36241	Pacemaker/defibrillator lead	CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A0203	Defective Device		
									A040101	Fracture		
									A070101	Failure to Capture		
									A070909	Oversensing		
									A072202	Low Impedance		

							F23	Unexpected Medical Intervention	A0203	Defective Device		
									A040101	Fracture		
									A070101	Failure to Capture		
									A070909	Oversensing		
									A072202	Low Impedance		
7907	2022	All other reportable incidents	35852	Single-chamber implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	27/03/2022	27/03/2022
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	27/03/2022	27/03/2022
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	27/03/2022	27/03/2022
7908	2022	All other reportable incidents	17882	Physiologic-monitoring defibrillation system	CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A0719	Unexpected Shutdown		
							F27	Problem Identified During Non-Clinical Procedure	A0719	Unexpected Shutdown		
7909	2022	All other reportable incidents	35852	Single-chamber implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem	29/09/2022	29/09/2022
									A070504	Premature Discharge of Battery	29/09/2022	29/09/2022
							F08	Hospitalization or Prolonged Hospitalization	A0705	Battery Problem	29/09/2022	29/09/2022
									A070504	Premature Discharge of Battery	29/09/2022	29/09/2022
							F1903	Device Explantation	A0705	Battery Problem	29/09/2022	29/09/2022
									A070504	Premature Discharge of Battery	29/09/2022	29/09/2022
7910	2022	All other reportable incidents	36241	Pacemaker/defibrillator lead	CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A040101	Fracture		
									A070101	Failure to Capture		
									A072201	High Impedance		
							F08	Hospitalization or Prolonged Hospitalization	A040101	Fracture		
									A070101	Failure to Capture		
									A072201	High Impedance		
							F1905	Device Revision or Replacement	A040101	Fracture		
									A070101	Failure to Capture		
									A072201	High Impedance		
7911	2022		47265	Dual-chamber implantable pacemaker, rate-responsive	CT141	Electrodes/Leads and associated devices	E0601	Arrhythmia	A24	No Apparent Device Problem		
							F02	Death	A24	No Apparent Device Problem		
					CT145	Cardiovascular devices	E0601	Arrhythmia	A24	No Apparent Device Problem		
							F02	Death	A24	No Apparent Device Problem		
					CT1684	Cardiac pacemakers and associated devices	E0601	Arrhythmia	A24	No Apparent Device Problem		
							F02	Death	A24	No Apparent Device Problem		
					CT1685	Cardiac pacemakers	E0601	Arrhythmia	A24	No Apparent Device Problem		
							F02	Death	A24	No Apparent Device Problem		
					CT233	Surgical	E0601	Arrhythmia	A24	No Apparent Device Problem		
							F02	Death	A24	No Apparent Device Problem		
					CT2406	Implantable	E0601	Arrhythmia	A24	No Apparent Device Problem		
							F02	Death	A24	No Apparent Device Problem		
					CT2407	Active implantable	E0601	Arrhythmia	A24	No Apparent Device Problem		
							F02	Death	A24	No Apparent Device Problem		
					CT257	Implantable cardiac pacemakers	E0601	Arrhythmia	A24	No Apparent Device Problem		
							F02	Death	A24	No Apparent Device Problem		
					CT2727	Active implant devices/ancillaries	E0601	Arrhythmia	A24	No Apparent Device Problem		
							F02	Death	A24	No Apparent Device Problem		
					CT2750	Implantable cardiac electrical stimulators and associated devices	E0601	Arrhythmia	A24	No Apparent Device Problem		
							F02	Death	A24	No Apparent Device Problem		
					CT2866	Implantable cardiac electrical stimulators	E0601	Arrhythmia	A24	No Apparent Device Problem		
							F02	Death	A24	No Apparent Device Problem		
					CT2985	Active	E0601	Arrhythmia	A24	No Apparent Device Problem		
							F02	Death	A24	No Apparent Device Problem		
					CT3001	Implantable metal	E0601	Arrhythmia	A24	No Apparent Device Problem		
							F02	Death	A24	No Apparent Device Problem		
					CT3002	Implantable synthetic polymer	E0601	Arrhythmia	A24	No Apparent Device Problem		
							F02	Death	A24	No Apparent Device Problem		
					CT3006	Implantable inorganic material	E0601	Arrhythmia	A24	No Apparent Device Problem		
							F02	Death	A24	No Apparent Device Problem		
					CT314	Electrically powered	E0601	Arrhythmia	A24	No Apparent Device Problem		
							F02	Death	A24	No Apparent Device Problem		
					CT321	Long-term surgical invasive	E0601	Arrhythmia	A24	No Apparent Device Problem		
							F02	Death	A24	No Apparent Device Problem		
					CT449	Generators and associated devices	E0601	Arrhythmia	A24	No Apparent Device Problem		
							F02	Death	A24	No Apparent Device Problem		
					CT981	Single use	E0601	Arrhythmia	A24	No Apparent Device Problem		
							F02	Death	A24	No Apparent Device Problem		
					CT983	Surgical invasive	E0601	Arrhythmia	A24	No Apparent Device Problem		
							F02	Death	A24	No Apparent Device Problem		
7912	2022	All other reportable incidents	47265	Dual-chamber implantable pacemaker, rate-responsive	CT141	Electrodes/Leads and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	16/09/2022	16/09/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	16/09/2022	16/09/2022
					CT145	Cardiovascular devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	16/09/2022	16/09/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	16/09/2022	16/09/2022
					CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	16/09/2022	16/09/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	16/09/2022	16/09/2022
					CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	16/09/2022	16/09/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	16/09/2022	16/09/2022
					CT233	Surgical	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	16/09/2022	16/09/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	16/09/2022	16/09/2022
					CT2406	Implantable	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	16/09/2022	16/09/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	16/09/2022	16/09/2022
					CT2407	Active implantable	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	16/09/2022	16/09/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	16/09/2022	16/09/2022
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	16/09/2022	16/09/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	16/09/2022	16/09/2022
					CT2727	Active implant devices/ancillaries	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	16/09/2022	16/09/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	16/09/2022	16/09/2022
					CT2750	Implantable cardiac electrical stimulators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	16/09/2022	16/09/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	16/09/2022	16/09/2022
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	16/09/2022	16/09/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	16/09/2022	16/09/2022
					CT2985	Active	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	16/09/2022	16/09/2022

							F1901	Additional Surgery	A24	No Apparent Device Problem	16/09/2022	16/09/2022
					CT3001	Implantable metal	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	16/09/2022	16/09/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	16/09/2022	16/09/2022
					CT3002	Implantable synthetic polymer	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	16/09/2022	16/09/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	16/09/2022	16/09/2022
					CT3006	Implantable inorganic material	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	16/09/2022	16/09/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	16/09/2022	16/09/2022
					CT314	Electrically powered	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	16/09/2022	16/09/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	16/09/2022	16/09/2022
					CT321	Long-term surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	16/09/2022	16/09/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	16/09/2022	16/09/2022
					CT449	Generators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	16/09/2022	16/09/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	16/09/2022	16/09/2022
					CT981	Single use	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	16/09/2022	16/09/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	16/09/2022	16/09/2022
					CT983	Surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	16/09/2022	16/09/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	16/09/2022	16/09/2022
7913	2022		47265	Dual-chamber implantable pacemaker, rate-responsive	CT141	Electrodes/Leads and associated devices	E0601	Arrhythmia	A0722	Impedance Problem	29/09/2022	29/09/2022
							F02	Death	A0722	Impedance Problem	29/09/2022	29/09/2022
					CT145	Cardiovascular devices	E0601	Arrhythmia	A0722	Impedance Problem	29/09/2022	29/09/2022
							F02	Death	A0722	Impedance Problem	29/09/2022	29/09/2022
					CT1684	Cardiac pacemakers and associated devices	E0601	Arrhythmia	A0722	Impedance Problem	29/09/2022	29/09/2022
							F02	Death	A0722	Impedance Problem	29/09/2022	29/09/2022
					CT1685	Cardiac pacemakers	E0601	Arrhythmia	A0722	Impedance Problem	29/09/2022	29/09/2022
							F02	Death	A0722	Impedance Problem	29/09/2022	29/09/2022
					CT233	Surgical	E0601	Arrhythmia	A0722	Impedance Problem	29/09/2022	29/09/2022
							F02	Death	A0722	Impedance Problem	29/09/2022	29/09/2022
					CT2406	Implantable	E0601	Arrhythmia	A0722	Impedance Problem	29/09/2022	29/09/2022
							F02	Death	A0722	Impedance Problem	29/09/2022	29/09/2022
					CT2407	Active implantable	E0601	Arrhythmia	A0722	Impedance Problem	29/09/2022	29/09/2022
							F02	Death	A0722	Impedance Problem	29/09/2022	29/09/2022
					CT257	Implantable cardiac pacemakers	E0601	Arrhythmia	A0722	Impedance Problem	29/09/2022	29/09/2022
							F02	Death	A0722	Impedance Problem	29/09/2022	29/09/2022
					CT2727	Active implant devices/ancillaries	E0601	Arrhythmia	A0722	Impedance Problem	29/09/2022	29/09/2022
							F02	Death	A0722	Impedance Problem	29/09/2022	29/09/2022
					CT2750	Implantable cardiac electrical stimulators and associated devices	E0601	Arrhythmia	A0722	Impedance Problem	29/09/2022	29/09/2022
							F02	Death	A0722	Impedance Problem	29/09/2022	29/09/2022
					CT2866	Implantable cardiac electrical stimulators	E0601	Arrhythmia	A0722	Impedance Problem	29/09/2022	29/09/2022
							F02	Death	A0722	Impedance Problem	29/09/2022	29/09/2022
					CT2985	Active	E0601	Arrhythmia	A0722	Impedance Problem	29/09/2022	29/09/2022
							F02	Death	A0722	Impedance Problem	29/09/2022	29/09/2022
					CT3001	Implantable metal	E0601	Arrhythmia	A0722	Impedance Problem	29/09/2022	29/09/2022
							F02	Death	A0722	Impedance Problem	29/09/2022	29/09/2022
					CT3002	Implantable synthetic polymer	E0601	Arrhythmia	A0722	Impedance Problem	29/09/2022	29/09/2022
							F02	Death	A0722	Impedance Problem	29/09/2022	29/09/2022
					CT3006	Implantable inorganic material	E0601	Arrhythmia	A0722	Impedance Problem	29/09/2022	29/09/2022
							F02	Death	A0722	Impedance Problem	29/09/2022	29/09/2022
					CT314	Electrically powered	E0601	Arrhythmia	A0722	Impedance Problem	29/09/2022	29/09/2022
							F02	Death	A0722	Impedance Problem	29/09/2022	29/09/2022
					CT321	Long-term surgical invasive	E0601	Arrhythmia	A0722	Impedance Problem	29/09/2022	29/09/2022
							F02	Death	A0722	Impedance Problem	29/09/2022	29/09/2022
					CT449	Generators and associated devices	E0601	Arrhythmia	A0722	Impedance Problem	29/09/2022	29/09/2022
							F02	Death	A0722	Impedance Problem	29/09/2022	29/09/2022
					CT981	Single use	E0601	Arrhythmia	A0722	Impedance Problem	29/09/2022	29/09/2022
							F02	Death	A0722	Impedance Problem	29/09/2022	29/09/2022
					CT983	Surgical invasive	E0601	Arrhythmia	A0722	Impedance Problem	29/09/2022	29/09/2022
							F02	Death	A0722	Impedance Problem	29/09/2022	29/09/2022
7914	2022	All other reportable incidents	47265	Dual-chamber implantable pacemaker, rate-responsive	CT141	Electrodes/Leads and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	21/09/2022	21/09/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	21/09/2022	21/09/2022
					CT145	Cardiovascular devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	21/09/2022	21/09/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	21/09/2022	21/09/2022
					CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	21/09/2022	21/09/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	21/09/2022	21/09/2022
					CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	21/09/2022	21/09/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	21/09/2022	21/09/2022
					CT233	Surgical	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	21/09/2022	21/09/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	21/09/2022	21/09/2022
					CT2406	Implantable	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	21/09/2022	21/09/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	21/09/2022	21/09/2022
					CT2407	Active implantable	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	21/09/2022	21/09/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	21/09/2022	21/09/2022
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	21/09/2022	21/09/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	21/09/2022	21/09/2022
					CT2727	Active implant devices/ancillaries	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	21/09/2022	21/09/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	21/09/2022	21/09/2022
					CT2750	Implantable cardiac electrical stimulators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	21/09/2022	21/09/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	21/09/2022	21/09/2022
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	21/09/2022	21/09/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	21/09/2022	21/09/2022
					CT2985	Active	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	21/09/2022	21/09/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	21/09/2022	21/09/2022
					CT3001	Implantable metal	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	21/09/2022	21/09/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	21/09/2022	21/09/2022
					CT3002	Implantable synthetic polymer	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	21/09/2022	21/09/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	21/09/2022	21/09/2022
					CT3006	Implantable inorganic material	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	21/09/2022	21/09/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	21/09/2022	21/09/2022
					CT314	Electrically powered	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	21/09/2022	21/09/2022

							F1901	Additional Surgery	A24	No Apparent Device Problem	21/09/2022	21/09/2022
					CT321	Long-term surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	21/09/2022	21/09/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	21/09/2022	21/09/2022
					CT449	Generators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	21/09/2022	21/09/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	21/09/2022	21/09/2022
					CT981	Single use	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	21/09/2022	21/09/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	21/09/2022	21/09/2022
					CT983	Surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	21/09/2022	21/09/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	21/09/2022	21/09/2022
7915	2022	All other reportable incidents	47265	Dual-chamber implantable pacemaker, rate-responsive	CT141	Electrodes/Leads and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	21/09/2022	21/09/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	21/09/2022	21/09/2022
					CT145	Cardiovascular devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	21/09/2022	21/09/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	21/09/2022	21/09/2022
					CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	21/09/2022	21/09/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	21/09/2022	21/09/2022
					CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	21/09/2022	21/09/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	21/09/2022	21/09/2022
					CT233	Surgical	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	21/09/2022	21/09/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	21/09/2022	21/09/2022
					CT2406	Implantable	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	21/09/2022	21/09/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	21/09/2022	21/09/2022
					CT2407	Active implantable	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	21/09/2022	21/09/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	21/09/2022	21/09/2022
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	21/09/2022	21/09/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	21/09/2022	21/09/2022
					CT2727	Active implant devices/ancillaries	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	21/09/2022	21/09/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	21/09/2022	21/09/2022
					CT2750	Implantable cardiac electrical stimulators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	21/09/2022	21/09/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	21/09/2022	21/09/2022
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	21/09/2022	21/09/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	21/09/2022	21/09/2022
					CT2985	Active	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	21/09/2022	21/09/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	21/09/2022	21/09/2022
					CT3001	Implantable metal	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	21/09/2022	21/09/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	21/09/2022	21/09/2022
					CT3002	Implantable synthetic polymer	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	21/09/2022	21/09/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	21/09/2022	21/09/2022
					CT3006	Implantable inorganic material	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	21/09/2022	21/09/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	21/09/2022	21/09/2022
					CT314	Electrically powered	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	21/09/2022	21/09/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	21/09/2022	21/09/2022
					CT321	Long-term surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	21/09/2022	21/09/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	21/09/2022	21/09/2022
					CT449	Generators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	21/09/2022	21/09/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	21/09/2022	21/09/2022
					CT981	Single use	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	21/09/2022	21/09/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	21/09/2022	21/09/2022
					CT983	Surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	21/09/2022	21/09/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	21/09/2022	21/09/2022
7916	2022	All other reportable incidents	47265	Dual-chamber implantable pacemaker, rate-responsive	CT141	Electrodes/Leads and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	24/09/2022	24/09/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	24/09/2022	24/09/2022
					CT145	Cardiovascular devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	24/09/2022	24/09/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	24/09/2022	24/09/2022
					CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	24/09/2022	24/09/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	24/09/2022	24/09/2022
					CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	24/09/2022	24/09/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	24/09/2022	24/09/2022
					CT233	Surgical	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	24/09/2022	24/09/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	24/09/2022	24/09/2022
					CT2406	Implantable	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	24/09/2022	24/09/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	24/09/2022	24/09/2022
					CT2407	Active implantable	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	24/09/2022	24/09/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	24/09/2022	24/09/2022
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	24/09/2022	24/09/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	24/09/2022	24/09/2022
					CT2727	Active implant devices/ancillaries	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	24/09/2022	24/09/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	24/09/2022	24/09/2022
					CT2750	Implantable cardiac electrical stimulators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	24/09/2022	24/09/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	24/09/2022	24/09/2022
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	24/09/2022	24/09/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	24/09/2022	24/09/2022
					CT2985	Active	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	24/09/2022	24/09/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	24/09/2022	24/09/2022
					CT3001	Implantable metal	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	24/09/2022	24/09/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	24/09/2022	24/09/2022
					CT3002	Implantable synthetic polymer	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	24/09/2022	24/09/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	24/09/2022	24/09/2022
					CT3006	Implantable inorganic material	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	24/09/2022	24/09/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	24/09/2022	24/09/2022
					CT314	Electrically powered	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	24/09/2022	24/09/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	24/09/2022	24/09/2022
					CT321	Long-term surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	24/09/2022	24/09/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	24/09/2022	24/09/2022
					CT449	Generators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	24/09/2022	24/09/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	24/09/2022	24/09/2022
					CT981	Single use	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	24/09/2022	24/09/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	24/09/2022	24/09/2022
					CT983	Surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	24/09/2022	24/09/2022

							F1901	Additional Surgery	A24	No Apparent Device Problem	24/09/2022	24/09/2022
7917	2022	All other reportable incidents	36241	Pacemaker/defibrillator lead	CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070909	Oversensing		
							F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A070909	Oversensing		
7918	2022	All other reportable incidents	35223	Endocardial/interventricular septal pacing lead	CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070909	Oversensing		
							F22	Unexpected Diagnostic Intervention	A070909	Oversensing		
7919	2022	All other reportable incidents	35223	Endocardial/interventricular septal pacing lead	CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070909	Oversensing		
							F1901	Additional Surgery	A070909	Oversensing		
7920	2022	All other reportable incidents	37265	Dual-chamber implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E0601	Arrhythmia	A071201	Failure to Convert Rhythm	17/02/2022	17/02/2022
							E2403	No Clinical Signs, Symptoms or Conditions	A071201	Failure to Convert Rhythm	17/02/2022	17/02/2022
							F1901	Additional Surgery	A071201	Failure to Convert Rhythm	17/02/2022	17/02/2022
7921	2022	All other reportable incidents	35852	Single-chamber implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem	11/11/2022	11/11/2022
									A070504	Premature Discharge of Battery	11/11/2022	11/11/2022
							F08	Hospitalization or Prolonged Hospitalization	A0705	Battery Problem	11/11/2022	11/11/2022
									A070504	Premature Discharge of Battery	11/11/2022	11/11/2022
							F1905	Device Revision or Replacement	A0705	Battery Problem	11/11/2022	11/11/2022
									A070504	Premature Discharge of Battery	11/11/2022	11/11/2022
7922	2022	All other reportable incidents	35852	Single-chamber implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
7923	2022	All other reportable incidents	47265	Dual-chamber implantable pacemaker, rate-responsive	CT141	Electrodes/Leads and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070101	Failure to Capture		
							F26	No Health Consequences or Impact	A070101	Failure to Capture		
					CT145	Cardiovascular devices	E2403	No Clinical Signs, Symptoms or Conditions	A070101	Failure to Capture		
							F26	No Health Consequences or Impact	A070101	Failure to Capture		
					CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070101	Failure to Capture		
							F26	No Health Consequences or Impact	A070101	Failure to Capture		
					CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A070101	Failure to Capture		
							F26	No Health Consequences or Impact	A070101	Failure to Capture		
					CT233	Surgical	E2403	No Clinical Signs, Symptoms or Conditions	A070101	Failure to Capture		
							F26	No Health Consequences or Impact	A070101	Failure to Capture		
					CT2406	Implantable	E2403	No Clinical Signs, Symptoms or Conditions	A070101	Failure to Capture		
							F26	No Health Consequences or Impact	A070101	Failure to Capture		
					CT2407	Active implantable	E2403	No Clinical Signs, Symptoms or Conditions	A070101	Failure to Capture		
							F26	No Health Consequences or Impact	A070101	Failure to Capture		
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A070101	Failure to Capture		
							F26	No Health Consequences or Impact	A070101	Failure to Capture		
					CT2727	Active implant devices/ancillaries	E2403	No Clinical Signs, Symptoms or Conditions	A070101	Failure to Capture		
							F26	No Health Consequences or Impact	A070101	Failure to Capture		
					CT2750	Implantable cardiac electrical stimulators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070101	Failure to Capture		
							F26	No Health Consequences or Impact	A070101	Failure to Capture		
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A070101	Failure to Capture		
							F26	No Health Consequences or Impact	A070101	Failure to Capture		
					CT2985	Active	E2403	No Clinical Signs, Symptoms or Conditions	A070101	Failure to Capture		
							F26	No Health Consequences or Impact	A070101	Failure to Capture		
					CT3001	Implantable metal	E2403	No Clinical Signs, Symptoms or Conditions	A070101	Failure to Capture		
							F26	No Health Consequences or Impact	A070101	Failure to Capture		
					CT3002	Implantable synthetic polymer	E2403	No Clinical Signs, Symptoms or Conditions	A070101	Failure to Capture		
							F26	No Health Consequences or Impact	A070101	Failure to Capture		
					CT3006	Implantable inorganic material	E2403	No Clinical Signs, Symptoms or Conditions	A070101	Failure to Capture		
							F26	No Health Consequences or Impact	A070101	Failure to Capture		
					CT314	Electrically powered	E2403	No Clinical Signs, Symptoms or Conditions	A070101	Failure to Capture		
							F26	No Health Consequences or Impact	A070101	Failure to Capture		
					CT321	Long-term surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A070101	Failure to Capture		
							F26	No Health Consequences or Impact	A070101	Failure to Capture		
					CT449	Generators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070101	Failure to Capture		
							F26	No Health Consequences or Impact	A070101	Failure to Capture		
					CT981	Single use	E2403	No Clinical Signs, Symptoms or Conditions	A070101	Failure to Capture		
							F26	No Health Consequences or Impact	A070101	Failure to Capture		
					CT983	Surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A070101	Failure to Capture		
							F26	No Health Consequences or Impact	A070101	Failure to Capture		
7924	2022	All other reportable incidents	47270	Cardiac resynchronization therapy implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	09/02/2022	09/02/2022
							F1901	Additional Surgery	A070504	Premature Discharge of Battery	09/02/2022	09/02/2022
7925	2022	All other reportable incidents	47270	Cardiac resynchronization therapy implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A072201	High Impedance		
							F26	No Health Consequences or Impact	A072201	High Impedance		
7926	2022	All other reportable incidents	47265	Dual-chamber implantable pacemaker, rate-responsive	CT141	Electrodes/Leads and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem		
							F26	No Health Consequences or Impact	A24	No Apparent Device Problem		
					CT145	Cardiovascular devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem		
							F26	No Health Consequences or Impact	A24	No Apparent Device Problem		
					CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem		
							F26	No Health Consequences or Impact	A24	No Apparent Device Problem		
					CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem		
							F26	No Health Consequences or Impact	A24	No Apparent Device Problem		
					CT233	Surgical	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem		
							F26	No Health Consequences or Impact	A24	No Apparent Device Problem		
					CT2406	Implantable	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem		
							F26	No Health Consequences or Impact	A24	No Apparent Device Problem		
					CT2407	Active implantable	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem		
							F26	No Health Consequences or Impact	A24	No Apparent Device Problem		
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem		
							F26	No Health Consequences or Impact	A24	No Apparent Device Problem		
					CT2727	Active implant devices/ancillaries	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem		
							F26	No Health Consequences or Impact	A24	No Apparent Device Problem		
					CT2750	Implantable cardiac electrical stimulators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem		
							F26	No Health Consequences or Impact	A24	No Apparent Device Problem		
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem		
							F26	No Health Consequences or Impact	A24	No Apparent Device Problem		
					CT2985	Active	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem		
							F26	No Health Consequences or Impact	A24	No Apparent Device Problem		
					CT3001	Implantable metal	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem		
							F26	No Health Consequences or Impact	A24	No Apparent Device Problem		

					CT3002	Implantable synthetic polymer	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem		
							F26	No Health Consequences or Impact	A24	No Apparent Device Problem		
					CT3006	Implantable inorganic material	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem		
							F26	No Health Consequences or Impact	A24	No Apparent Device Problem		
					CT314	Electrically powered	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem		
							F26	No Health Consequences or Impact	A24	No Apparent Device Problem		
					CT321	Long-term surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem		
							F26	No Health Consequences or Impact	A24	No Apparent Device Problem		
					CT449	Generators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem		
							F26	No Health Consequences or Impact	A24	No Apparent Device Problem		
					CT981	Single use	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem		
							F26	No Health Consequences or Impact	A24	No Apparent Device Problem		
					CT983	Surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem		
							F26	No Health Consequences or Impact	A24	No Apparent Device Problem		
7927	2022	All other reportable incidents	17882	Physiologic-monitoring defibrillation system	CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A090209	Visual Prompts Will not Clear		
							F27	Problem Identified During Non-Clinical Procedure	A090209	Visual Prompts Will not Clear		
7928	2022	All other reportable incidents	17882	Physiologic-monitoring defibrillation system	CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A090812	Missing Test Results		
							F27	Problem Identified During Non-Clinical Procedure	A090812	Missing Test Results		
7929	2022	Unanticipated serious deterioration in state of health	37265	Dual-chamber implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E060109	Tachycardia	A0709	Device Sensing Problem		
									A071301	Failure to Deliver Shock/Stimulation		
							E0602	Cardiac Arrest	A0709	Device Sensing Problem		
									A071301	Failure to Deliver Shock/Stimulation		
							F08	Hospitalization or Prolonged Hospitalization	A0709	Device Sensing Problem		
									A071301	Failure to Deliver Shock/Stimulation		
							F1001	Absence of Treatment	A0709	Device Sensing Problem		
									A071301	Failure to Deliver Shock/Stimulation		
							F1203	Life Threatening Illness or Injury	A0709	Device Sensing Problem		
									A071301	Failure to Deliver Shock/Stimulation		
							F23	Unexpected Medical Intervention	A0709	Device Sensing Problem		
									A071301	Failure to Deliver Shock/Stimulation		
							F2306	Resuscitation	A0709	Device Sensing Problem		
									A071301	Failure to Deliver Shock/Stimulation		
7930	2022	All other reportable incidents	35852	Single-chamber implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
									A071102	Failure to Interrogate		
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery		
									A071102	Failure to Interrogate		
							F12	Serious Injury/ Illness/ Impairment	A070504	Premature Discharge of Battery		
									A071102	Failure to Interrogate		
							F1903	Device Explantation	A070504	Premature Discharge of Battery		
									A071102	Failure to Interrogate		
7931	2022	All other reportable incidents	37265	Dual-chamber implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
							F27	Problem Identified During Non-Clinical Procedure	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
7932	2022		47265	Dual-chamber implantable pacemaker, rate-responsive	CT141	Electrodes/Leads and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	23/08/2022	23/08/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	23/08/2022	23/08/2022
					CT145	Cardiovascular devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	23/08/2022	23/08/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	23/08/2022	23/08/2022
					CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	23/08/2022	23/08/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	23/08/2022	23/08/2022
					CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	23/08/2022	23/08/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	23/08/2022	23/08/2022
					CT233	Surgical	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	23/08/2022	23/08/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	23/08/2022	23/08/2022
					CT2406	Implantable	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	23/08/2022	23/08/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	23/08/2022	23/08/2022
					CT2407	Active implantable	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	23/08/2022	23/08/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	23/08/2022	23/08/2022
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	23/08/2022	23/08/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	23/08/2022	23/08/2022
					CT2727	Active implant devices/ancillaries	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	23/08/2022	23/08/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	23/08/2022	23/08/2022
					CT2750	Implantable cardiac electrical stimulators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	23/08/2022	23/08/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	23/08/2022	23/08/2022
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	23/08/2022	23/08/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	23/08/2022	23/08/2022
					CT2985	Active	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	23/08/2022	23/08/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	23/08/2022	23/08/2022
					CT3001	Implantable metal	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	23/08/2022	23/08/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	23/08/2022	23/08/2022
					CT3002	Implantable synthetic polymer	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	23/08/2022	23/08/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	23/08/2022	23/08/2022
					CT3006	Implantable inorganic material	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	23/08/2022	23/08/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	23/08/2022	23/08/2022
					CT314	Electrically powered	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	23/08/2022	23/08/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	23/08/2022	23/08/2022
					CT321	Long-term surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	23/08/2022	23/08/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	23/08/2022	23/08/2022
					CT449	Generators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	23/08/2022	23/08/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	23/08/2022	23/08/2022
					CT981	Single use	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	23/08/2022	23/08/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	23/08/2022	23/08/2022
					CT983	Surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	23/08/2022	23/08/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	23/08/2022	23/08/2022
7933	2022	All other reportable incidents	35852	Single-chamber implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem	09/03/2022	09/03/2022
									A070504	Premature Discharge of Battery	09/03/2022	09/03/2022
							F08	Hospitalization or Prolonged Hospitalization	A0705	Battery Problem	09/03/2022	09/03/2022
									A070504	Premature Discharge of Battery	09/03/2022	09/03/2022

						F1905	Device Revision or Replacement	A0705	Battery Problem	09/03/2022	09/03/2022
								A070504	Premature Discharge of Battery	09/03/2022	09/03/2022
7934	2022	All other reportable incidents	47265	Dual-chamber implantable pacemaker, rate-responsive	CT141	Electrodes/Leads and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	24/09/2022	24/09/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	24/09/2022	24/09/2022
					CT145	Cardiovascular devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	24/09/2022	24/09/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	24/09/2022	24/09/2022
					CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	24/09/2022	24/09/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	24/09/2022	24/09/2022
					CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	24/09/2022	24/09/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	24/09/2022	24/09/2022
					CT233	Surgical	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	24/09/2022	24/09/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	24/09/2022	24/09/2022
					CT2406	Implantable	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	24/09/2022	24/09/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	24/09/2022	24/09/2022
					CT2407	Active implantable	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	24/09/2022	24/09/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	24/09/2022	24/09/2022
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	24/09/2022	24/09/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	24/09/2022	24/09/2022
					CT2727	Active implant devices/ancillaries	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	24/09/2022	24/09/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	24/09/2022	24/09/2022
					CT2750	Implantable cardiac electrical stimulators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	24/09/2022	24/09/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	24/09/2022	24/09/2022
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	24/09/2022	24/09/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	24/09/2022	24/09/2022
					CT2985	Active	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	24/09/2022	24/09/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	24/09/2022	24/09/2022
					CT3001	Implantable metal	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	24/09/2022	24/09/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	24/09/2022	24/09/2022
					CT3002	Implantable synthetic polymer	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	24/09/2022	24/09/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	24/09/2022	24/09/2022
					CT3006	Implantable inorganic material	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	24/09/2022	24/09/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	24/09/2022	24/09/2022
					CT314	Electrically powered	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	24/09/2022	24/09/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	24/09/2022	24/09/2022
					CT321	Long-term surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	24/09/2022	24/09/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	24/09/2022	24/09/2022
					CT449	Generators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	24/09/2022	24/09/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	24/09/2022	24/09/2022
					CT981	Single use	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	24/09/2022	24/09/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	24/09/2022	24/09/2022
					CT983	Surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	24/09/2022	24/09/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	24/09/2022	24/09/2022
7935	2022	All other reportable incidents	47270	Cardiac resynchronization therapy implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A072201	High Impedance		
							F26	No Health Consequences or Impact	A072201	High Impedance		
7936	2022		47265	Dual-chamber implantable pacemaker, rate-responsive	CT141	Electrodes/Leads and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	24/09/2022	24/09/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	24/09/2022	24/09/2022
					CT145	Cardiovascular devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	24/09/2022	24/09/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	24/09/2022	24/09/2022
					CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	24/09/2022	24/09/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	24/09/2022	24/09/2022
					CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	24/09/2022	24/09/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	24/09/2022	24/09/2022
					CT233	Surgical	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	24/09/2022	24/09/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	24/09/2022	24/09/2022
					CT2406	Implantable	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	24/09/2022	24/09/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	24/09/2022	24/09/2022
					CT2407	Active implantable	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	24/09/2022	24/09/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	24/09/2022	24/09/2022
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	24/09/2022	24/09/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	24/09/2022	24/09/2022
					CT2727	Active implant devices/ancillaries	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	24/09/2022	24/09/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	24/09/2022	24/09/2022
					CT2750	Implantable cardiac electrical stimulators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	24/09/2022	24/09/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	24/09/2022	24/09/2022
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	24/09/2022	24/09/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	24/09/2022	24/09/2022
					CT2985	Active	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	24/09/2022	24/09/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	24/09/2022	24/09/2022
					CT3001	Implantable metal	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	24/09/2022	24/09/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	24/09/2022	24/09/2022
					CT3002	Implantable synthetic polymer	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	24/09/2022	24/09/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	24/09/2022	24/09/2022
					CT3006	Implantable inorganic material	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	24/09/2022	24/09/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	24/09/2022	24/09/2022
					CT314	Electrically powered	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	24/09/2022	24/09/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	24/09/2022	24/09/2022
					CT321	Long-term surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	24/09/2022	24/09/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	24/09/2022	24/09/2022
					CT449	Generators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	24/09/2022	24/09/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	24/09/2022	24/09/2022
					CT981	Single use	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	24/09/2022	24/09/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	24/09/2022	24/09/2022
					CT983	Surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	24/09/2022	24/09/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	24/09/2022	24/09/2022
7937	2022	All other reportable incidents	47265	Dual-chamber implantable pacemaker, rate-responsive	CT141	Electrodes/Leads and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	24/09/2022	24/09/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	24/09/2022	24/09/2022
					CT145	Cardiovascular devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	24/09/2022	24/09/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	24/09/2022	24/09/2022

					CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	24/09/2022	24/09/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	24/09/2022	24/09/2022
					CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	24/09/2022	24/09/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	24/09/2022	24/09/2022
					CT233	Surgical	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	24/09/2022	24/09/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	24/09/2022	24/09/2022
					CT2406	Implantable	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	24/09/2022	24/09/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	24/09/2022	24/09/2022
					CT2407	Active implantable	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	24/09/2022	24/09/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	24/09/2022	24/09/2022
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	24/09/2022	24/09/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	24/09/2022	24/09/2022
					CT2727	Active implant devices/ancillaries	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	24/09/2022	24/09/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	24/09/2022	24/09/2022
					CT2750	Implantable cardiac electrical stimulators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	24/09/2022	24/09/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	24/09/2022	24/09/2022
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	24/09/2022	24/09/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	24/09/2022	24/09/2022
					CT2985	Active	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	24/09/2022	24/09/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	24/09/2022	24/09/2022
					CT3001	Implantable metal	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	24/09/2022	24/09/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	24/09/2022	24/09/2022
					CT3002	Implantable synthetic polymer	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	24/09/2022	24/09/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	24/09/2022	24/09/2022
					CT3006	Implantable inorganic material	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	24/09/2022	24/09/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	24/09/2022	24/09/2022
					CT314	Electrically powered	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	24/09/2022	24/09/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	24/09/2022	24/09/2022
					CT321	Long-term surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	24/09/2022	24/09/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	24/09/2022	24/09/2022
					CT449	Generators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	24/09/2022	24/09/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	24/09/2022	24/09/2022
					CT981	Single use	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	24/09/2022	24/09/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	24/09/2022	24/09/2022
					CT983	Surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	24/09/2022	24/09/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	24/09/2022	24/09/2022
7938	2022	All other reportable incidents	47265	Dual-chamber implantable pacemaker, rate-responsive	CT141	Electrodes/Leads and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	16/09/2022	16/09/2022
							F08	Hospitalization or Prolonged Hospitalization	A24	No Apparent Device Problem	16/09/2022	16/09/2022
							F1905	Device Revision or Replacement	A24	No Apparent Device Problem	16/09/2022	16/09/2022
					CT145	Cardiovascular devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	16/09/2022	16/09/2022
							F08	Hospitalization or Prolonged Hospitalization	A24	No Apparent Device Problem	16/09/2022	16/09/2022
							F1905	Device Revision or Replacement	A24	No Apparent Device Problem	16/09/2022	16/09/2022
					CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	16/09/2022	16/09/2022
							F08	Hospitalization or Prolonged Hospitalization	A24	No Apparent Device Problem	16/09/2022	16/09/2022
							F1905	Device Revision or Replacement	A24	No Apparent Device Problem	16/09/2022	16/09/2022
					CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	16/09/2022	16/09/2022
							F08	Hospitalization or Prolonged Hospitalization	A24	No Apparent Device Problem	16/09/2022	16/09/2022
							F1905	Device Revision or Replacement	A24	No Apparent Device Problem	16/09/2022	16/09/2022
					CT233	Surgical	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	16/09/2022	16/09/2022
							F08	Hospitalization or Prolonged Hospitalization	A24	No Apparent Device Problem	16/09/2022	16/09/2022
							F1905	Device Revision or Replacement	A24	No Apparent Device Problem	16/09/2022	16/09/2022
					CT2406	Implantable	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	16/09/2022	16/09/2022
							F08	Hospitalization or Prolonged Hospitalization	A24	No Apparent Device Problem	16/09/2022	16/09/2022
							F1905	Device Revision or Replacement	A24	No Apparent Device Problem	16/09/2022	16/09/2022
					CT2407	Active implantable	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	16/09/2022	16/09/2022
							F08	Hospitalization or Prolonged Hospitalization	A24	No Apparent Device Problem	16/09/2022	16/09/2022
							F1905	Device Revision or Replacement	A24	No Apparent Device Problem	16/09/2022	16/09/2022
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	16/09/2022	16/09/2022
							F08	Hospitalization or Prolonged Hospitalization	A24	No Apparent Device Problem	16/09/2022	16/09/2022
							F1905	Device Revision or Replacement	A24	No Apparent Device Problem	16/09/2022	16/09/2022
					CT2727	Active implant devices/ancillaries	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	16/09/2022	16/09/2022
							F08	Hospitalization or Prolonged Hospitalization	A24	No Apparent Device Problem	16/09/2022	16/09/2022
							F1905	Device Revision or Replacement	A24	No Apparent Device Problem	16/09/2022	16/09/2022
					CT2750	Implantable cardiac electrical stimulators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	16/09/2022	16/09/2022
							F08	Hospitalization or Prolonged Hospitalization	A24	No Apparent Device Problem	16/09/2022	16/09/2022
							F1905	Device Revision or Replacement	A24	No Apparent Device Problem	16/09/2022	16/09/2022
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	16/09/2022	16/09/2022
							F08	Hospitalization or Prolonged Hospitalization	A24	No Apparent Device Problem	16/09/2022	16/09/2022
							F1905	Device Revision or Replacement	A24	No Apparent Device Problem	16/09/2022	16/09/2022
					CT2985	Active	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	16/09/2022	16/09/2022
							F08	Hospitalization or Prolonged Hospitalization	A24	No Apparent Device Problem	16/09/2022	16/09/2022
							F1905	Device Revision or Replacement	A24	No Apparent Device Problem	16/09/2022	16/09/2022
					CT3001	Implantable metal	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	16/09/2022	16/09/2022
							F08	Hospitalization or Prolonged Hospitalization	A24	No Apparent Device Problem	16/09/2022	16/09/2022
							F1905	Device Revision or Replacement	A24	No Apparent Device Problem	16/09/2022	16/09/2022
					CT3002	Implantable synthetic polymer	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	16/09/2022	16/09/2022
							F08	Hospitalization or Prolonged Hospitalization	A24	No Apparent Device Problem	16/09/2022	16/09/2022
							F1905	Device Revision or Replacement	A24	No Apparent Device Problem	16/09/2022	16/09/2022
					CT3006	Implantable inorganic material	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	16/09/2022	16/09/2022
							F08	Hospitalization or Prolonged Hospitalization	A24	No Apparent Device Problem	16/09/2022	16/09/2022
							F1905	Device Revision or Replacement	A24	No Apparent Device Problem	16/09/2022	16/09/2022
					CT314	Electrically powered	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	16/09/2022	16/09/2022
							F08	Hospitalization or Prolonged Hospitalization	A24	No Apparent Device Problem	16/09/2022	16/09/2022
							F1905	Device Revision or Replacement	A24	No Apparent Device Problem	16/09/2022	16/09/2022
					CT321	Long-term surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	16/09/2022	16/09/2022
							F08	Hospitalization or Prolonged Hospitalization	A24	No Apparent Device Problem	16/09/2022	16/09/2022
							F1905	Device Revision or Replacement	A24	No Apparent Device Problem	16/09/2022	16/09/2022
					CT449	Generators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	16/09/2022	16/09/2022

							F08	Hospitalization or Prolonged Hospitalization	A24	No Apparent Device Problem	16/09/2022	16/09/2022
							F1905	Device Revision or Replacement	A24	No Apparent Device Problem	16/09/2022	16/09/2022
					CT981	Single use	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	16/09/2022	16/09/2022
							F08	Hospitalization or Prolonged Hospitalization	A24	No Apparent Device Problem	16/09/2022	16/09/2022
							F1905	Device Revision or Replacement	A24	No Apparent Device Problem	16/09/2022	16/09/2022
					CT983	Surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	16/09/2022	16/09/2022
							F08	Hospitalization or Prolonged Hospitalization	A24	No Apparent Device Problem	16/09/2022	16/09/2022
							F1905	Device Revision or Replacement	A24	No Apparent Device Problem	16/09/2022	16/09/2022
7939	2022	All other reportable incidents	47267	Single-chamber implantable pacemaker, rate-responsive	CT141	Electrodes/Leads and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT145	Cardiovascular devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT233	Surgical	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT2406	Implantable	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT2407	Active implantable	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT2727	Active implant devices/ancillaries	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT2750	Implantable cardiac electrical stimulators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT2985	Active	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT3001	Implantable metal	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT3002	Implantable synthetic polymer	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT3006	Implantable inorganic material	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT314	Electrically powered	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT321	Long-term surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT449	Generators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT981	Single use	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT983	Surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
7940	2022	All other reportable incidents	47270	Cardiac resynchronization therapy implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A071301	Failure to Deliver Shock/Stimulation		
									A090805	Incorrect Measurement		
							F26	No Health Consequences or Impact	A071301	Failure to Deliver Shock/Stimulation		
									A090805	Incorrect Measurement		
7941	2022	All other reportable incidents	47265	Dual-chamber implantable pacemaker, rate-responsive	CT141	Electrodes/Leads and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	24/09/2022	24/09/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	24/09/2022	24/09/2022
					CT145	Cardiovascular devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	24/09/2022	24/09/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	24/09/2022	24/09/2022
					CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	24/09/2022	24/09/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	24/09/2022	24/09/2022
					CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	24/09/2022	24/09/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	24/09/2022	24/09/2022
					CT233	Surgical	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	24/09/2022	24/09/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	24/09/2022	24/09/2022
					CT2406	Implantable	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	24/09/2022	24/09/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	24/09/2022	24/09/2022
					CT2407	Active implantable	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	24/09/2022	24/09/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	24/09/2022	24/09/2022
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	24/09/2022	24/09/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	24/09/2022	24/09/2022
					CT2727	Active implant devices/ancillaries	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	24/09/2022	24/09/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	24/09/2022	24/09/2022
					CT2750	Implantable cardiac electrical stimulators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	24/09/2022	24/09/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	24/09/2022	24/09/2022
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	24/09/2022	24/09/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	24/09/2022	24/09/2022
					CT2985	Active	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	24/09/2022	24/09/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	24/09/2022	24/09/2022
					CT3001	Implantable metal	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	24/09/2022	24/09/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	24/09/2022	24/09/2022
					CT3002	Implantable synthetic polymer	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	24/09/2022	24/09/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	24/09/2022	24/09/2022
					CT3006	Implantable inorganic material	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	24/09/2022	24/09/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	24/09/2022	24/09/2022
					CT314	Electrically powered	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	24/09/2022	24/09/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	24/09/2022	24/09/2022
					CT321	Long-term surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	24/09/2022	24/09/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	24/09/2022	24/09/2022
					CT449	Generators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	24/09/2022	24/09/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	24/09/2022	24/09/2022

					CT981	Single use	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	24/09/2022	24/09/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	24/09/2022	24/09/2022
					CT983	Surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	24/09/2022	24/09/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	24/09/2022	24/09/2022
7942	2022	All other reportable incidents	47265	Dual-chamber implantable pacemaker, rate-responsive	CT141	Electrodes/Leads and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
					CT145	Cardiovascular devices	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
					CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
					CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
					CT233	Surgical	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
					CT2406	Implantable	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
					CT2407	Active implantable	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
					CT2727	Active implant devices/ancillaries	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
					CT2750	Implantable cardiac electrical stimulators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
					CT2985	Active	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
					CT3001	Implantable metal	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
					CT3002	Implantable synthetic polymer	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
					CT3006	Implantable inorganic material	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
					CT314	Electrically powered	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
					CT321	Long-term surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
					CT449	Generators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
					CT981	Single use	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
					CT983	Surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
7943	2022	All other reportable incidents	36241	Pacemaker/defibrillator lead	CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A0203	Defective Device		
									A072202	Low Impedance		
							F08	Hospitalization or Prolonged Hospitalization	A0203	Defective Device		
									A072202	Low Impedance		

							F1905	Device Revision or Replacement	A0203	Defective Device			
									A072202	Low Impedance			
7944	2022	All other reportable incidents	47265	Dual-chamber implantable pacemaker, rate-responsive	CT141	Electrodes/Leads and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	30/08/2022	30/08/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	30/08/2022	30/08/2022
					CT145	Cardiovascular devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	30/08/2022	30/08/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	30/08/2022	30/08/2022
					CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	30/08/2022	30/08/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	30/08/2022	30/08/2022
					CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	30/08/2022	30/08/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	30/08/2022	30/08/2022
					CT233	Surgical	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	30/08/2022	30/08/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	30/08/2022	30/08/2022
					CT2406	Implantable	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	30/08/2022	30/08/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	30/08/2022	30/08/2022
					CT2407	Active implantable	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	30/08/2022	30/08/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	30/08/2022	30/08/2022
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	30/08/2022	30/08/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	30/08/2022	30/08/2022
					CT2727	Active implant devices/ancillaries	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	30/08/2022	30/08/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	30/08/2022	30/08/2022
					CT2750	Implantable cardiac electrical stimulators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	30/08/2022	30/08/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	30/08/2022	30/08/2022
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	30/08/2022	30/08/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	30/08/2022	30/08/2022
					CT2985	Active	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	30/08/2022	30/08/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	30/08/2022	30/08/2022
					CT3001	Implantable metal	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	30/08/2022	30/08/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	30/08/2022	30/08/2022
					CT3002	Implantable synthetic polymer	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	30/08/2022	30/08/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	30/08/2022	30/08/2022
					CT3006	Implantable inorganic material	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	30/08/2022	30/08/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	30/08/2022	30/08/2022
					CT314	Electrically powered	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	30/08/2022	30/08/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	30/08/2022	30/08/2022
					CT321	Long-term surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	30/08/2022	30/08/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	30/08/2022	30/08/2022
					CT449	Generators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	30/08/2022	30/08/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	30/08/2022	30/08/2022
					CT981	Single use	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	30/08/2022	30/08/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	30/08/2022	30/08/2022
					CT983	Surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	30/08/2022	30/08/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	30/08/2022	30/08/2022
7945	2022	All other reportable incidents	47265	Dual-chamber implantable pacemaker, rate-responsive	CT141	Electrodes/Leads and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	01/10/2022	01/10/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	01/10/2022	01/10/2022
					CT145	Cardiovascular devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	01/10/2022	01/10/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	01/10/2022	01/10/2022
					CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	01/10/2022	01/10/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	01/10/2022	01/10/2022
					CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	01/10/2022	01/10/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	01/10/2022	01/10/2022
					CT233	Surgical	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	01/10/2022	01/10/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	01/10/2022	01/10/2022
					CT2406	Implantable	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	01/10/2022	01/10/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	01/10/2022	01/10/2022
					CT2407	Active implantable	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	01/10/2022	01/10/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	01/10/2022	01/10/2022
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	01/10/2022	01/10/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	01/10/2022	01/10/2022
					CT2727	Active implant devices/ancillaries	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	01/10/2022	01/10/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	01/10/2022	01/10/2022
					CT2750	Implantable cardiac electrical stimulators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	01/10/2022	01/10/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	01/10/2022	01/10/2022
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	01/10/2022	01/10/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	01/10/2022	01/10/2022
					CT2985	Active	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	01/10/2022	01/10/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	01/10/2022	01/10/2022
					CT3001	Implantable metal	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	01/10/2022	01/10/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	01/10/2022	01/10/2022
					CT3002	Implantable synthetic polymer	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	01/10/2022	01/10/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	01/10/2022	01/10/2022
					CT3006	Implantable inorganic material	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	01/10/2022	01/10/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	01/10/2022	01/10/2022
					CT314	Electrically powered	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	01/10/2022	01/10/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	01/10/2022	01/10/2022
					CT321	Long-term surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	01/10/2022	01/10/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	01/10/2022	01/10/2022
					CT449	Generators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	01/10/2022	01/10/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	01/10/2022	01/10/2022
					CT981	Single use	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	01/10/2022	01/10/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	01/10/2022	01/10/2022
					CT983	Surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	01/10/2022	01/10/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	01/10/2022	01/10/2022
7946	2022		47265	Dual-chamber implantable pacemaker, rate-responsive	CT141	Electrodes/Leads and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	23/08/2022	23/08/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	23/08/2022	23/08/2022
					CT145	Cardiovascular devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	23/08/2022	23/08/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	23/08/2022	23/08/2022
					CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	23/08/2022	23/08/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	23/08/2022	23/08/2022

					CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	23/08/2022	23/08/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	23/08/2022	23/08/2022
					CT233	Surgical	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	23/08/2022	23/08/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	23/08/2022	23/08/2022
					CT2406	Implantable	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	23/08/2022	23/08/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	23/08/2022	23/08/2022
					CT2407	Active implantable	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	23/08/2022	23/08/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	23/08/2022	23/08/2022
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	23/08/2022	23/08/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	23/08/2022	23/08/2022
					CT2727	Active implant devices/ancillaries	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	23/08/2022	23/08/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	23/08/2022	23/08/2022
					CT2750	Implantable cardiac electrical stimulators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	23/08/2022	23/08/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	23/08/2022	23/08/2022
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	23/08/2022	23/08/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	23/08/2022	23/08/2022
					CT2985	Active	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	23/08/2022	23/08/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	23/08/2022	23/08/2022
					CT3001	Implantable metal	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	23/08/2022	23/08/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	23/08/2022	23/08/2022
					CT3002	Implantable synthetic polymer	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	23/08/2022	23/08/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	23/08/2022	23/08/2022
					CT3006	Implantable inorganic material	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	23/08/2022	23/08/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	23/08/2022	23/08/2022
					CT314	Electrically powered	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	23/08/2022	23/08/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	23/08/2022	23/08/2022
					CT321	Long-term surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	23/08/2022	23/08/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	23/08/2022	23/08/2022
					CT449	Generators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	23/08/2022	23/08/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	23/08/2022	23/08/2022
					CT981	Single use	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	23/08/2022	23/08/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	23/08/2022	23/08/2022
					CT983	Surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	23/08/2022	23/08/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	23/08/2022	23/08/2022
7947	2022		47265	Dual-chamber implantable pacemaker, rate-responsive	CT141	Electrodes/Leads and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	28/09/2022	28/09/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	28/09/2022	28/09/2022
					CT145	Cardiovascular devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	28/09/2022	28/09/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	28/09/2022	28/09/2022
					CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	28/09/2022	28/09/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	28/09/2022	28/09/2022
					CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	28/09/2022	28/09/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	28/09/2022	28/09/2022
					CT233	Surgical	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	28/09/2022	28/09/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	28/09/2022	28/09/2022
					CT2406	Implantable	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	28/09/2022	28/09/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	28/09/2022	28/09/2022
					CT2407	Active implantable	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	28/09/2022	28/09/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	28/09/2022	28/09/2022
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	28/09/2022	28/09/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	28/09/2022	28/09/2022
					CT2727	Active implant devices/ancillaries	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	28/09/2022	28/09/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	28/09/2022	28/09/2022
					CT2750	Implantable cardiac electrical stimulators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	28/09/2022	28/09/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	28/09/2022	28/09/2022
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	28/09/2022	28/09/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	28/09/2022	28/09/2022
					CT2985	Active	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	28/09/2022	28/09/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	28/09/2022	28/09/2022
					CT3001	Implantable metal	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	28/09/2022	28/09/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	28/09/2022	28/09/2022
					CT3002	Implantable synthetic polymer	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	28/09/2022	28/09/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	28/09/2022	28/09/2022
					CT3006	Implantable inorganic material	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	28/09/2022	28/09/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	28/09/2022	28/09/2022
					CT314	Electrically powered	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	28/09/2022	28/09/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	28/09/2022	28/09/2022
					CT321	Long-term surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	28/09/2022	28/09/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	28/09/2022	28/09/2022
					CT449	Generators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	28/09/2022	28/09/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	28/09/2022	28/09/2022
					CT981	Single use	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	28/09/2022	28/09/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	28/09/2022	28/09/2022
					CT983	Surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	28/09/2022	28/09/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	28/09/2022	28/09/2022
7948	2022		47265	Dual-chamber implantable pacemaker, rate-responsive	CT141	Electrodes/Leads and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	28/09/2022	28/09/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	28/09/2022	28/09/2022
					CT145	Cardiovascular devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	28/09/2022	28/09/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	28/09/2022	28/09/2022
					CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	28/09/2022	28/09/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	28/09/2022	28/09/2022
					CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	28/09/2022	28/09/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	28/09/2022	28/09/2022
					CT233	Surgical	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	28/09/2022	28/09/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	28/09/2022	28/09/2022
					CT2406	Implantable	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	28/09/2022	28/09/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	28/09/2022	28/09/2022
					CT2407	Active implantable	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	28/09/2022	28/09/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	28/09/2022	28/09/2022

					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	28/09/2022	28/09/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	28/09/2022	28/09/2022
					CT2727	Active implant devices/ancillaries	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	28/09/2022	28/09/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	28/09/2022	28/09/2022
					CT2750	Implantable cardiac electrical stimulators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	28/09/2022	28/09/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	28/09/2022	28/09/2022
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	28/09/2022	28/09/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	28/09/2022	28/09/2022
					CT2985	Active	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	28/09/2022	28/09/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	28/09/2022	28/09/2022
					CT3001	Implantable metal	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	28/09/2022	28/09/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	28/09/2022	28/09/2022
					CT3002	Implantable synthetic polymer	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	28/09/2022	28/09/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	28/09/2022	28/09/2022
					CT3006	Implantable inorganic material	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	28/09/2022	28/09/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	28/09/2022	28/09/2022
					CT314	Electrically powered	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	28/09/2022	28/09/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	28/09/2022	28/09/2022
					CT321	Long-term surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	28/09/2022	28/09/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	28/09/2022	28/09/2022
					CT449	Generators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	28/09/2022	28/09/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	28/09/2022	28/09/2022
					CT981	Single use	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	28/09/2022	28/09/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	28/09/2022	28/09/2022
					CT983	Surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	28/09/2022	28/09/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	28/09/2022	28/09/2022
7949	2022	All other reportable incidents	47265	Dual-chamber implantable pacemaker, rate-responsive	CT141	Electrodes/Leads and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	29/09/2022	29/09/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	29/09/2022	29/09/2022
					CT145	Cardiovascular devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	29/09/2022	29/09/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	29/09/2022	29/09/2022
					CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	29/09/2022	29/09/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	29/09/2022	29/09/2022
					CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	29/09/2022	29/09/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	29/09/2022	29/09/2022
					CT233	Surgical	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	29/09/2022	29/09/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	29/09/2022	29/09/2022
					CT2406	Implantable	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	29/09/2022	29/09/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	29/09/2022	29/09/2022
					CT2407	Active implantable	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	29/09/2022	29/09/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	29/09/2022	29/09/2022
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	29/09/2022	29/09/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	29/09/2022	29/09/2022
					CT2727	Active implant devices/ancillaries	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	29/09/2022	29/09/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	29/09/2022	29/09/2022
					CT2750	Implantable cardiac electrical stimulators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	29/09/2022	29/09/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	29/09/2022	29/09/2022
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	29/09/2022	29/09/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	29/09/2022	29/09/2022
					CT2985	Active	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	29/09/2022	29/09/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	29/09/2022	29/09/2022
					CT3001	Implantable metal	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	29/09/2022	29/09/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	29/09/2022	29/09/2022
					CT3002	Implantable synthetic polymer	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	29/09/2022	29/09/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	29/09/2022	29/09/2022
					CT3006	Implantable inorganic material	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	29/09/2022	29/09/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	29/09/2022	29/09/2022
					CT314	Electrically powered	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	29/09/2022	29/09/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	29/09/2022	29/09/2022
					CT321	Long-term surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	29/09/2022	29/09/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	29/09/2022	29/09/2022
					CT449	Generators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	29/09/2022	29/09/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	29/09/2022	29/09/2022
					CT981	Single use	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	29/09/2022	29/09/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	29/09/2022	29/09/2022
					CT983	Surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	29/09/2022	29/09/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	29/09/2022	29/09/2022
7950	2022		47265	Dual-chamber implantable pacemaker, rate-responsive	CT141	Electrodes/Leads and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	01/10/2022	01/10/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	01/10/2022	01/10/2022
					CT145	Cardiovascular devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	01/10/2022	01/10/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	01/10/2022	01/10/2022
					CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	01/10/2022	01/10/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	01/10/2022	01/10/2022
					CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	01/10/2022	01/10/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	01/10/2022	01/10/2022
					CT233	Surgical	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	01/10/2022	01/10/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	01/10/2022	01/10/2022
					CT2406	Implantable	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	01/10/2022	01/10/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	01/10/2022	01/10/2022
					CT2407	Active implantable	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	01/10/2022	01/10/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	01/10/2022	01/10/2022
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	01/10/2022	01/10/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	01/10/2022	01/10/2022
					CT2727	Active implant devices/ancillaries	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	01/10/2022	01/10/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	01/10/2022	01/10/2022
					CT2750	Implantable cardiac electrical stimulators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	01/10/2022	01/10/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	01/10/2022	01/10/2022
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	01/10/2022	01/10/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	01/10/2022	01/10/2022

					CT2985	Active	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	01/10/2022	01/10/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	01/10/2022	01/10/2022
					CT3001	Implantable metal	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	01/10/2022	01/10/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	01/10/2022	01/10/2022
					CT3002	Implantable synthetic polymer	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	01/10/2022	01/10/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	01/10/2022	01/10/2022
					CT3006	Implantable inorganic material	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	01/10/2022	01/10/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	01/10/2022	01/10/2022
					CT314	Electrically powered	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	01/10/2022	01/10/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	01/10/2022	01/10/2022
					CT321	Long-term surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	01/10/2022	01/10/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	01/10/2022	01/10/2022
					CT449	Generators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	01/10/2022	01/10/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	01/10/2022	01/10/2022
					CT981	Single use	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	01/10/2022	01/10/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	01/10/2022	01/10/2022
					CT983	Surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	01/10/2022	01/10/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	01/10/2022	01/10/2022
7951	2022		47265	Dual-chamber implantable pacemaker, rate-responsive	CT141	Electrodes/Leads and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	01/10/2022	01/10/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	01/10/2022	01/10/2022
					CT145	Cardiovascular devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	01/10/2022	01/10/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	01/10/2022	01/10/2022
					CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	01/10/2022	01/10/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	01/10/2022	01/10/2022
					CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	01/10/2022	01/10/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	01/10/2022	01/10/2022
					CT233	Surgical	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	01/10/2022	01/10/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	01/10/2022	01/10/2022
					CT2406	Implantable	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	01/10/2022	01/10/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	01/10/2022	01/10/2022
					CT2407	Active implantable	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	01/10/2022	01/10/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	01/10/2022	01/10/2022
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	01/10/2022	01/10/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	01/10/2022	01/10/2022
					CT2727	Active implant devices/ancillaries	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	01/10/2022	01/10/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	01/10/2022	01/10/2022
					CT2750	Implantable cardiac electrical stimulators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	01/10/2022	01/10/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	01/10/2022	01/10/2022
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	01/10/2022	01/10/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	01/10/2022	01/10/2022
					CT2985	Active	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	01/10/2022	01/10/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	01/10/2022	01/10/2022
					CT3001	Implantable metal	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	01/10/2022	01/10/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	01/10/2022	01/10/2022
					CT3002	Implantable synthetic polymer	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	01/10/2022	01/10/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	01/10/2022	01/10/2022
					CT3006	Implantable inorganic material	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	01/10/2022	01/10/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	01/10/2022	01/10/2022
					CT314	Electrically powered	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	01/10/2022	01/10/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	01/10/2022	01/10/2022
					CT321	Long-term surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	01/10/2022	01/10/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	01/10/2022	01/10/2022
					CT449	Generators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	01/10/2022	01/10/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	01/10/2022	01/10/2022
					CT981	Single use	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	01/10/2022	01/10/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	01/10/2022	01/10/2022
					CT983	Surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	01/10/2022	01/10/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	01/10/2022	01/10/2022
7952	2022	All other reportable incidents	37265	Dual-chamber implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E2104	Electric Shock	A0712	Pacing Problem		
									A071302	Inappropriate/Inadequate Shock/Stimulation		
							E2403	No Clinical Signs, Symptoms or Conditions	A0712	Pacing Problem		
									A071302	Inappropriate/Inadequate Shock/Stimulation		
							F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A0712	Pacing Problem		
									A071302	Inappropriate/Inadequate Shock/Stimulation		
7953	2022	All other reportable incidents	47265	Dual-chamber implantable pacemaker, rate-responsive	CT141	Electrodes/Leads and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
					CT145	Cardiovascular devices	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
					CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
					CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
					CT233	Surgical	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
					CT2406	Implantable	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		

					CT2407	Active implantable	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
					CT2727	Active implant devices/ancillaries	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
					CT2750	Implantable cardiac electrical stimulators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
					CT2985	Active	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
					CT3001	Implantable metal	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
					CT3002	Implantable synthetic polymer	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
					CT3006	Implantable inorganic material	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
					CT314	Electrically powered	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
					CT321	Long-term surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
					CT449	Generators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
					CT981	Single use	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
					CT983	Surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
7954	2022	All other reportable incidents	17882	Physiologic-monitoring defibrillation system	CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070803	Failure to Power Up		
							F27	Problem Identified During Non-Clinical Procedure	A070803	Failure to Power Up		
7955	2022	All other reportable incidents	36241	Pacemaker/defibrillator lead	CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A072201	High Impedance		
							F26	No Health Consequences or Impact	A072201	High Impedance		
7956	2022	All other reportable incidents	17882	Physiologic-monitoring defibrillation system	CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070802	Intermittent Loss of Power		
							F27	Problem Identified During Non-Clinical Procedure	A070802	Intermittent Loss of Power		
7957	2022	All other reportable incidents	47265	Dual-chamber implantable pacemaker, rate-responsive	CT141	Electrodes/Leads and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT145	Cardiovascular devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT233	Surgical	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT2406	Implantable	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT2407	Active implantable	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT2727	Active implant devices/ancillaries	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT2750	Implantable cardiac electrical stimulators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT2985	Active	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT3001	Implantable metal	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		

					CT3002	Implantable synthetic polymer	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT3006	Implantable inorganic material	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT314	Electrically powered	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT321	Long-term surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT449	Generators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT981	Single use	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT983	Surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
7958	2022	All other reportable incidents	36241	Pacemaker/defibrillator lead	CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F27	Problem Identified During Non-Clinical Procedure	A070504	Premature Discharge of Battery		
7959	2022	All other reportable incidents	47265	Dual-chamber implantable pacemaker, rate-responsive	CT141	Electrodes/Leads and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem	28/04/2022	28/04/2022
									A070504	Premature Discharge of Battery	28/04/2022	28/04/2022
							F08	Hospitalization or Prolonged Hospitalization	A0705	Battery Problem	28/04/2022	28/04/2022
									A070504	Premature Discharge of Battery	28/04/2022	28/04/2022
							F1905	Device Revision or Replacement	A0705	Battery Problem	28/04/2022	28/04/2022
									A070504	Premature Discharge of Battery	28/04/2022	28/04/2022
					CT145	Cardiovascular devices	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem	28/04/2022	28/04/2022
									A070504	Premature Discharge of Battery	28/04/2022	28/04/2022
							F08	Hospitalization or Prolonged Hospitalization	A0705	Battery Problem	28/04/2022	28/04/2022
									A070504	Premature Discharge of Battery	28/04/2022	28/04/2022
							F1905	Device Revision or Replacement	A0705	Battery Problem	28/04/2022	28/04/2022
									A070504	Premature Discharge of Battery	28/04/2022	28/04/2022
					CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem	28/04/2022	28/04/2022
									A070504	Premature Discharge of Battery	28/04/2022	28/04/2022
							F08	Hospitalization or Prolonged Hospitalization	A0705	Battery Problem	28/04/2022	28/04/2022
									A070504	Premature Discharge of Battery	28/04/2022	28/04/2022
							F1905	Device Revision or Replacement	A0705	Battery Problem	28/04/2022	28/04/2022
									A070504	Premature Discharge of Battery	28/04/2022	28/04/2022
					CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem	28/04/2022	28/04/2022
									A070504	Premature Discharge of Battery	28/04/2022	28/04/2022
							F08	Hospitalization or Prolonged Hospitalization	A0705	Battery Problem	28/04/2022	28/04/2022
									A070504	Premature Discharge of Battery	28/04/2022	28/04/2022
							F1905	Device Revision or Replacement	A0705	Battery Problem	28/04/2022	28/04/2022
									A070504	Premature Discharge of Battery	28/04/2022	28/04/2022
					CT233	Surgical	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem	28/04/2022	28/04/2022
									A070504	Premature Discharge of Battery	28/04/2022	28/04/2022
							F08	Hospitalization or Prolonged Hospitalization	A0705	Battery Problem	28/04/2022	28/04/2022
									A070504	Premature Discharge of Battery	28/04/2022	28/04/2022
							F1905	Device Revision or Replacement	A0705	Battery Problem	28/04/2022	28/04/2022
									A070504	Premature Discharge of Battery	28/04/2022	28/04/2022
					CT2406	Implantable	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem	28/04/2022	28/04/2022
									A070504	Premature Discharge of Battery	28/04/2022	28/04/2022
							F08	Hospitalization or Prolonged Hospitalization	A0705	Battery Problem	28/04/2022	28/04/2022
									A070504	Premature Discharge of Battery	28/04/2022	28/04/2022
							F1905	Device Revision or Replacement	A0705	Battery Problem	28/04/2022	28/04/2022
									A070504	Premature Discharge of Battery	28/04/2022	28/04/2022
					CT2407	Active implantable	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem	28/04/2022	28/04/2022
									A070504	Premature Discharge of Battery	28/04/2022	28/04/2022
							F08	Hospitalization or Prolonged Hospitalization	A0705	Battery Problem	28/04/2022	28/04/2022
									A070504	Premature Discharge of Battery	28/04/2022	28/04/2022
							F1905	Device Revision or Replacement	A0705	Battery Problem	28/04/2022	28/04/2022
									A070504	Premature Discharge of Battery	28/04/2022	28/04/2022
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem	28/04/2022	28/04/2022
									A070504	Premature Discharge of Battery	28/04/2022	28/04/2022
							F08	Hospitalization or Prolonged Hospitalization	A0705	Battery Problem	28/04/2022	28/04/2022
									A070504	Premature Discharge of Battery	28/04/2022	28/04/2022
							F1905	Device Revision or Replacement	A0705	Battery Problem	28/04/2022	28/04/2022
									A070504	Premature Discharge of Battery	28/04/2022	28/04/2022
					CT2727	Active implant devices/ancillaries	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem	28/04/2022	28/04/2022
									A070504	Premature Discharge of Battery	28/04/2022	28/04/2022
							F08	Hospitalization or Prolonged Hospitalization	A0705	Battery Problem	28/04/2022	28/04/2022
									A070504	Premature Discharge of Battery	28/04/2022	28/04/2022
							F1905	Device Revision or Replacement	A0705	Battery Problem	28/04/2022	28/04/2022
									A070504	Premature Discharge of Battery	28/04/2022	28/04/2022
					CT2750	Implantable cardiac electrical stimulators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem	28/04/2022	28/04/2022
									A070504	Premature Discharge of Battery	28/04/2022	28/04/2022
							F08	Hospitalization or Prolonged Hospitalization	A0705	Battery Problem	28/04/2022	28/04/2022
									A070504	Premature Discharge of Battery	28/04/2022	28/04/2022
							F1905	Device Revision or Replacement	A0705	Battery Problem	28/04/2022	28/04/2022
									A070504	Premature Discharge of Battery	28/04/2022	28/04/2022
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem	28/04/2022	28/04/2022
									A070504	Premature Discharge of Battery	28/04/2022	28/04/2022
							F08	Hospitalization or Prolonged Hospitalization	A0705	Battery Problem	28/04/2022	28/04/2022
									A070504	Premature Discharge of Battery	28/04/2022	28/04/2022
							F1905	Device Revision or Replacement	A0705	Battery Problem	28/04/2022	28/04/2022
									A070504	Premature Discharge of Battery	28/04/2022	28/04/2022
					CT2985	Active	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem	28/04/2022	28/04/2022
									A070504	Premature Discharge of Battery	28/04/2022	28/04/2022
							F08	Hospitalization or Prolonged Hospitalization	A0705	Battery Problem	28/04/2022	28/04/2022
									A070504	Premature Discharge of Battery	28/04/2022	28/04/2022
							F1905	Device Revision or Replacement	A0705	Battery Problem	28/04/2022	28/04/2022
									A070504	Premature Discharge of Battery	28/04/2022	28/04/2022

					CT3001	Implantable metal	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem	28/04/2022	28/04/2022
									A070504	Premature Discharge of Battery	28/04/2022	28/04/2022
							F08	Hospitalization or Prolonged Hospitalization	A0705	Battery Problem	28/04/2022	28/04/2022
									A070504	Premature Discharge of Battery	28/04/2022	28/04/2022
							F1905	Device Revision or Replacement	A0705	Battery Problem	28/04/2022	28/04/2022
									A070504	Premature Discharge of Battery	28/04/2022	28/04/2022
					CT3002	Implantable synthetic polymer	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem	28/04/2022	28/04/2022
									A070504	Premature Discharge of Battery	28/04/2022	28/04/2022
							F08	Hospitalization or Prolonged Hospitalization	A0705	Battery Problem	28/04/2022	28/04/2022
									A070504	Premature Discharge of Battery	28/04/2022	28/04/2022
							F1905	Device Revision or Replacement	A0705	Battery Problem	28/04/2022	28/04/2022
									A070504	Premature Discharge of Battery	28/04/2022	28/04/2022
					CT3006	Implantable inorganic material	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem	28/04/2022	28/04/2022
									A070504	Premature Discharge of Battery	28/04/2022	28/04/2022
							F08	Hospitalization or Prolonged Hospitalization	A0705	Battery Problem	28/04/2022	28/04/2022
									A070504	Premature Discharge of Battery	28/04/2022	28/04/2022
							F1905	Device Revision or Replacement	A0705	Battery Problem	28/04/2022	28/04/2022
									A070504	Premature Discharge of Battery	28/04/2022	28/04/2022
					CT314	Electrically powered	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem	28/04/2022	28/04/2022
									A070504	Premature Discharge of Battery	28/04/2022	28/04/2022
							F08	Hospitalization or Prolonged Hospitalization	A0705	Battery Problem	28/04/2022	28/04/2022
									A070504	Premature Discharge of Battery	28/04/2022	28/04/2022
							F1905	Device Revision or Replacement	A0705	Battery Problem	28/04/2022	28/04/2022
									A070504	Premature Discharge of Battery	28/04/2022	28/04/2022
					CT321	Long-term surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem	28/04/2022	28/04/2022
									A070504	Premature Discharge of Battery	28/04/2022	28/04/2022
							F08	Hospitalization or Prolonged Hospitalization	A0705	Battery Problem	28/04/2022	28/04/2022
									A070504	Premature Discharge of Battery	28/04/2022	28/04/2022
							F1905	Device Revision or Replacement	A0705	Battery Problem	28/04/2022	28/04/2022
									A070504	Premature Discharge of Battery	28/04/2022	28/04/2022
					CT449	Generators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem	28/04/2022	28/04/2022
									A070504	Premature Discharge of Battery	28/04/2022	28/04/2022
							F08	Hospitalization or Prolonged Hospitalization	A0705	Battery Problem	28/04/2022	28/04/2022
									A070504	Premature Discharge of Battery	28/04/2022	28/04/2022
							F1905	Device Revision or Replacement	A0705	Battery Problem	28/04/2022	28/04/2022
									A070504	Premature Discharge of Battery	28/04/2022	28/04/2022
					CT981	Single use	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem	28/04/2022	28/04/2022
									A070504	Premature Discharge of Battery	28/04/2022	28/04/2022
							F08	Hospitalization or Prolonged Hospitalization	A0705	Battery Problem	28/04/2022	28/04/2022
									A070504	Premature Discharge of Battery	28/04/2022	28/04/2022
							F1905	Device Revision or Replacement	A0705	Battery Problem	28/04/2022	28/04/2022
									A070504	Premature Discharge of Battery	28/04/2022	28/04/2022
					CT983	Surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem	28/04/2022	28/04/2022
									A070504	Premature Discharge of Battery	28/04/2022	28/04/2022
							F08	Hospitalization or Prolonged Hospitalization	A0705	Battery Problem	28/04/2022	28/04/2022
									A070504	Premature Discharge of Battery	28/04/2022	28/04/2022
							F1905	Device Revision or Replacement	A0705	Battery Problem	28/04/2022	28/04/2022
									A070504	Premature Discharge of Battery	28/04/2022	28/04/2022
7960	2022	All other reportable incidents	36241	Pacemaker/defibrillator lead	CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A05	Mechanical Problem	08/11/2021	08/11/2021
							F1908	Prolonged Surgery	A05	Mechanical Problem	08/11/2021	08/11/2021
7961	2022	All other reportable incidents	47270	Cardiac resynchronization therapy implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A070909	Oversensing		
							F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A070909	Oversensing		
7962	2022		47265	Dual-chamber implantable pacemaker, rate-responsive	CT141	Electrodes/Leads and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	01/10/2022	01/10/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	01/10/2022	01/10/2022
					CT145	Cardiovascular devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	01/10/2022	01/10/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	01/10/2022	01/10/2022
					CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	01/10/2022	01/10/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	01/10/2022	01/10/2022
					CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	01/10/2022	01/10/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	01/10/2022	01/10/2022
					CT233	Surgical	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	01/10/2022	01/10/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	01/10/2022	01/10/2022
					CT2406	Implantable	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	01/10/2022	01/10/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	01/10/2022	01/10/2022
					CT2407	Active implantable	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	01/10/2022	01/10/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	01/10/2022	01/10/2022
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	01/10/2022	01/10/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	01/10/2022	01/10/2022
					CT2727	Active implant devices/ancillaries	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	01/10/2022	01/10/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	01/10/2022	01/10/2022
					CT2750	Implantable cardiac electrical stimulators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	01/10/2022	01/10/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	01/10/2022	01/10/2022
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	01/10/2022	01/10/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	01/10/2022	01/10/2022
					CT2985	Active	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	01/10/2022	01/10/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	01/10/2022	01/10/2022
					CT3001	Implantable metal	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	01/10/2022	01/10/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	01/10/2022	01/10/2022
					CT3002	Implantable synthetic polymer	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	01/10/2022	01/10/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	01/10/2022	01/10/2022
					CT3006	Implantable inorganic material	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	01/10/2022	01/10/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	01/10/2022	01/10/2022
					CT314	Electrically powered	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	01/10/2022	01/10/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	01/10/2022	01/10/2022
					CT321	Long-term surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	01/10/2022	01/10/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	01/10/2022	01/10/2022
					CT449	Generators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	01/10/2022	01/10/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	01/10/2022	01/10/2022

					CT981	Single use	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	01/10/2022	01/10/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	01/10/2022	01/10/2022
					CT983	Surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	01/10/2022	01/10/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	01/10/2022	01/10/2022
7963	2022	All other reportable incidents	36241	Pacemaker/defibrillator lead	CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070101	Failure to Capture		
									A072201	High Impedance		
							F2203	Imaging Required	A070101	Failure to Capture		
									A072201	High Impedance		
7964	2022		47265	Dual-chamber implantable pacemaker, rate-responsive	CT141	Electrodes/Leads and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	01/10/2022	01/10/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	01/10/2022	01/10/2022
					CT145	Cardiovascular devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	01/10/2022	01/10/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	01/10/2022	01/10/2022
					CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	01/10/2022	01/10/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	01/10/2022	01/10/2022
					CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	01/10/2022	01/10/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	01/10/2022	01/10/2022
					CT233	Surgical	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	01/10/2022	01/10/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	01/10/2022	01/10/2022
					CT2406	Implantable	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	01/10/2022	01/10/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	01/10/2022	01/10/2022
					CT2407	Active implantable	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	01/10/2022	01/10/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	01/10/2022	01/10/2022
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	01/10/2022	01/10/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	01/10/2022	01/10/2022
					CT2727	Active implant devices/ancillaries	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	01/10/2022	01/10/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	01/10/2022	01/10/2022
					CT2750	Implantable cardiac electrical stimulators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	01/10/2022	01/10/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	01/10/2022	01/10/2022
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	01/10/2022	01/10/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	01/10/2022	01/10/2022
					CT2985	Active	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	01/10/2022	01/10/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	01/10/2022	01/10/2022
					CT3001	Implantable metal	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	01/10/2022	01/10/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	01/10/2022	01/10/2022
					CT3002	Implantable synthetic polymer	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	01/10/2022	01/10/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	01/10/2022	01/10/2022
					CT3006	Implantable inorganic material	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	01/10/2022	01/10/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	01/10/2022	01/10/2022
					CT314	Electrically powered	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	01/10/2022	01/10/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	01/10/2022	01/10/2022
					CT321	Long-term surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	01/10/2022	01/10/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	01/10/2022	01/10/2022
					CT449	Generators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	01/10/2022	01/10/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	01/10/2022	01/10/2022
					CT981	Single use	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	01/10/2022	01/10/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	01/10/2022	01/10/2022
					CT983	Surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	01/10/2022	01/10/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	01/10/2022	01/10/2022
7965	2022	All other reportable incidents	47265	Dual-chamber implantable pacemaker, rate-responsive	CT141	Electrodes/Leads and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	29/09/2022	29/09/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	29/09/2022	29/09/2022
					CT145	Cardiovascular devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	29/09/2022	29/09/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	29/09/2022	29/09/2022
					CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	29/09/2022	29/09/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	29/09/2022	29/09/2022
					CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	29/09/2022	29/09/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	29/09/2022	29/09/2022
					CT233	Surgical	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	29/09/2022	29/09/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	29/09/2022	29/09/2022
					CT2406	Implantable	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	29/09/2022	29/09/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	29/09/2022	29/09/2022
					CT2407	Active implantable	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	29/09/2022	29/09/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	29/09/2022	29/09/2022
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	29/09/2022	29/09/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	29/09/2022	29/09/2022
					CT2727	Active implant devices/ancillaries	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	29/09/2022	29/09/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	29/09/2022	29/09/2022
					CT2750	Implantable cardiac electrical stimulators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	29/09/2022	29/09/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	29/09/2022	29/09/2022
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	29/09/2022	29/09/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	29/09/2022	29/09/2022
					CT2985	Active	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	29/09/2022	29/09/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	29/09/2022	29/09/2022
					CT3001	Implantable metal	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	29/09/2022	29/09/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	29/09/2022	29/09/2022
					CT3002	Implantable synthetic polymer	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	29/09/2022	29/09/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	29/09/2022	29/09/2022
					CT3006	Implantable inorganic material	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	29/09/2022	29/09/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	29/09/2022	29/09/2022
					CT314	Electrically powered	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	29/09/2022	29/09/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	29/09/2022	29/09/2022
					CT321	Long-term surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	29/09/2022	29/09/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	29/09/2022	29/09/2022
					CT449	Generators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	29/09/2022	29/09/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	29/09/2022	29/09/2022
					CT981	Single use	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	29/09/2022	29/09/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	29/09/2022	29/09/2022
					CT983	Surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	29/09/2022	29/09/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	29/09/2022	29/09/2022

7966	2022	All other reportable incidents	47265	Dual-chamber implantable pacemaker, rate-responsive	CT141	Electrodes/Leads and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	29/09/2022	29/09/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	29/09/2022	29/09/2022
					CT145	Cardiovascular devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	29/09/2022	29/09/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	29/09/2022	29/09/2022
					CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	29/09/2022	29/09/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	29/09/2022	29/09/2022
					CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	29/09/2022	29/09/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	29/09/2022	29/09/2022
					CT233	Surgical	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	29/09/2022	29/09/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	29/09/2022	29/09/2022
					CT2406	Implantable	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	29/09/2022	29/09/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	29/09/2022	29/09/2022
					CT2407	Active implantable	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	29/09/2022	29/09/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	29/09/2022	29/09/2022
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	29/09/2022	29/09/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	29/09/2022	29/09/2022
					CT2727	Active implant devices/ancillaries	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	29/09/2022	29/09/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	29/09/2022	29/09/2022
					CT2750	Implantable cardiac electrical stimulators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	29/09/2022	29/09/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	29/09/2022	29/09/2022
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	29/09/2022	29/09/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	29/09/2022	29/09/2022
					CT2985	Active	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	29/09/2022	29/09/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	29/09/2022	29/09/2022
					CT3001	Implantable metal	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	29/09/2022	29/09/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	29/09/2022	29/09/2022
					CT3002	Implantable synthetic polymer	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	29/09/2022	29/09/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	29/09/2022	29/09/2022
					CT3006	Implantable inorganic material	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	29/09/2022	29/09/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	29/09/2022	29/09/2022
					CT314	Electrically powered	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	29/09/2022	29/09/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	29/09/2022	29/09/2022
					CT321	Long-term surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	29/09/2022	29/09/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	29/09/2022	29/09/2022
					CT449	Generators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	29/09/2022	29/09/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	29/09/2022	29/09/2022
					CT981	Single use	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	29/09/2022	29/09/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	29/09/2022	29/09/2022
					CT983	Surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	29/09/2022	29/09/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	29/09/2022	29/09/2022
7967	2022	All other reportable incidents	35223	Endocardial/interventricular septal pacing lead	CT1684	Cardiac pacemakers and associated devices	E011903	Syncope/Fainting	A070909	Oversensing		
							E2403	No Clinical Signs, Symptoms or Conditions	A070909	Oversensing		
							F1901	Additional Surgery	A070909	Oversensing		
7968	2022	All other reportable incidents	35852	Single-chamber implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
									A0908	Incorrect, Inadequate or Imprecise Result or Readings		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
									A0908	Incorrect, Inadequate or Imprecise Result or Readings		
7969	2022	All other reportable incidents	35852	Single-chamber implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem	09/02/2022	09/02/2022
									A070504	Premature Discharge of Battery	09/02/2022	09/02/2022
							F08	Hospitalization or Prolonged Hospitalization	A0705	Battery Problem	09/02/2022	09/02/2022
									A070504	Premature Discharge of Battery	09/02/2022	09/02/2022
							F1905	Device Revision or Replacement	A0705	Battery Problem	09/02/2022	09/02/2022
									A070504	Premature Discharge of Battery	09/02/2022	09/02/2022
7970	2022	All other reportable incidents	36241	Pacemaker/defibrillator lead	CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A051201	Device Dislodged or Dislocated		
									A070909	Oversensing		
									A090801	Signal Artifact/Noise		
									A1301	Failure to Read Input Signal		
							F15	Recognized Device or Procedural Complication	A051201	Device Dislodged or Dislocated		
									A070909	Oversensing		
									A090801	Signal Artifact/Noise		
									A1301	Failure to Read Input Signal		
							F2203	Imaging Required	A051201	Device Dislodged or Dislocated		
									A070909	Oversensing		
									A090801	Signal Artifact/Noise		
									A1301	Failure to Read Input Signal		
7971	2022	All other reportable incidents	35852	Single-chamber implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F27	Problem Identified During Non-Clinical Procedure	A070504	Premature Discharge of Battery		
7972	2022	All other reportable incidents	35852	Single-chamber implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A072201	High Impedance		
							F26	No Health Consequences or Impact	A072201	High Impedance		
7973	2022	All other reportable incidents	36241	Pacemaker/defibrillator lead	CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A072201	High Impedance		
							F26	No Health Consequences or Impact	A072201	High Impedance		
7974	2022		47265	Dual-chamber implantable pacemaker, rate-responsive	CT141	Electrodes/Leads and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	01/10/2022	01/10/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	01/10/2022	01/10/2022
					CT145	Cardiovascular devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	01/10/2022	01/10/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	01/10/2022	01/10/2022
					CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	01/10/2022	01/10/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	01/10/2022	01/10/2022
					CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	01/10/2022	01/10/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	01/10/2022	01/10/2022
					CT233	Surgical	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	01/10/2022	01/10/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	01/10/2022	01/10/2022
					CT2406	Implantable	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	01/10/2022	01/10/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	01/10/2022	01/10/2022
					CT2407	Active implantable	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	01/10/2022	01/10/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	01/10/2022	01/10/2022
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	01/10/2022	01/10/2022

						F1901	Additional Surgery	A24	No Apparent Device Problem	01/10/2022	01/10/2022
				CT2727	Active implant devices/ancillaries	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	01/10/2022	01/10/2022
						F1901	Additional Surgery	A24	No Apparent Device Problem	01/10/2022	01/10/2022
				CT2750	Implantable cardiac electrical stimulators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	01/10/2022	01/10/2022
						F1901	Additional Surgery	A24	No Apparent Device Problem	01/10/2022	01/10/2022
				CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	01/10/2022	01/10/2022
						F1901	Additional Surgery	A24	No Apparent Device Problem	01/10/2022	01/10/2022
				CT2985	Active	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	01/10/2022	01/10/2022
						F1901	Additional Surgery	A24	No Apparent Device Problem	01/10/2022	01/10/2022
				CT3001	Implantable metal	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	01/10/2022	01/10/2022
						F1901	Additional Surgery	A24	No Apparent Device Problem	01/10/2022	01/10/2022
				CT3002	Implantable synthetic polymer	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	01/10/2022	01/10/2022
						F1901	Additional Surgery	A24	No Apparent Device Problem	01/10/2022	01/10/2022
				CT3006	Implantable inorganic material	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	01/10/2022	01/10/2022
						F1901	Additional Surgery	A24	No Apparent Device Problem	01/10/2022	01/10/2022
				CT314	Electrically powered	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	01/10/2022	01/10/2022
						F1901	Additional Surgery	A24	No Apparent Device Problem	01/10/2022	01/10/2022
				CT321	Long-term surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	01/10/2022	01/10/2022
						F1901	Additional Surgery	A24	No Apparent Device Problem	01/10/2022	01/10/2022
				CT449	Generators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	01/10/2022	01/10/2022
						F1901	Additional Surgery	A24	No Apparent Device Problem	01/10/2022	01/10/2022
				CT981	Single use	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	01/10/2022	01/10/2022
						F1901	Additional Surgery	A24	No Apparent Device Problem	01/10/2022	01/10/2022
				CT983	Surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	01/10/2022	01/10/2022
						F1901	Additional Surgery	A24	No Apparent Device Problem	01/10/2022	01/10/2022
7975	2022	All other reportable incidents	47265	Dual-chamber implantable pacemaker, rate-responsive	CT141	Electrodes/Leads and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	29/09/2022	29/09/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	29/09/2022	29/09/2022
				CT145	Cardiovascular devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	29/09/2022	29/09/2022
						F1901	Additional Surgery	A24	No Apparent Device Problem	29/09/2022	29/09/2022
				CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	29/09/2022	29/09/2022
						F1901	Additional Surgery	A24	No Apparent Device Problem	29/09/2022	29/09/2022
				CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	29/09/2022	29/09/2022
						F1901	Additional Surgery	A24	No Apparent Device Problem	29/09/2022	29/09/2022
				CT233	Surgical	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	29/09/2022	29/09/2022
						F1901	Additional Surgery	A24	No Apparent Device Problem	29/09/2022	29/09/2022
				CT2406	Implantable	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	29/09/2022	29/09/2022
						F1901	Additional Surgery	A24	No Apparent Device Problem	29/09/2022	29/09/2022
				CT2407	Active implantable	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	29/09/2022	29/09/2022
						F1901	Additional Surgery	A24	No Apparent Device Problem	29/09/2022	29/09/2022
				CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	29/09/2022	29/09/2022
						F1901	Additional Surgery	A24	No Apparent Device Problem	29/09/2022	29/09/2022
				CT2727	Active implant devices/ancillaries	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	29/09/2022	29/09/2022
						F1901	Additional Surgery	A24	No Apparent Device Problem	29/09/2022	29/09/2022
				CT2750	Implantable cardiac electrical stimulators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	29/09/2022	29/09/2022
						F1901	Additional Surgery	A24	No Apparent Device Problem	29/09/2022	29/09/2022
				CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	29/09/2022	29/09/2022
						F1901	Additional Surgery	A24	No Apparent Device Problem	29/09/2022	29/09/2022
				CT2985	Active	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	29/09/2022	29/09/2022
						F1901	Additional Surgery	A24	No Apparent Device Problem	29/09/2022	29/09/2022
				CT3001	Implantable metal	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	29/09/2022	29/09/2022
						F1901	Additional Surgery	A24	No Apparent Device Problem	29/09/2022	29/09/2022
				CT3002	Implantable synthetic polymer	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	29/09/2022	29/09/2022
						F1901	Additional Surgery	A24	No Apparent Device Problem	29/09/2022	29/09/2022
				CT3006	Implantable inorganic material	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	29/09/2022	29/09/2022
						F1901	Additional Surgery	A24	No Apparent Device Problem	29/09/2022	29/09/2022
				CT314	Electrically powered	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	29/09/2022	29/09/2022
						F1901	Additional Surgery	A24	No Apparent Device Problem	29/09/2022	29/09/2022
				CT321	Long-term surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	29/09/2022	29/09/2022
						F1901	Additional Surgery	A24	No Apparent Device Problem	29/09/2022	29/09/2022
				CT449	Generators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	29/09/2022	29/09/2022
						F1901	Additional Surgery	A24	No Apparent Device Problem	29/09/2022	29/09/2022
				CT981	Single use	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	29/09/2022	29/09/2022
						F1901	Additional Surgery	A24	No Apparent Device Problem	29/09/2022	29/09/2022
				CT983	Surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	29/09/2022	29/09/2022
						F1901	Additional Surgery	A24	No Apparent Device Problem	29/09/2022	29/09/2022
7976	2022	All other reportable incidents	47267	Single-chamber implantable pacemaker, rate-responsive	CT141	Electrodes/Leads and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
				CT145	Cardiovascular devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
						F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
				CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
						F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
				CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
						F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
				CT233	Surgical	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
						F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
				CT2406	Implantable	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
						F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
				CT2407	Active implantable	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
						F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
				CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
						F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
				CT2727	Active implant devices/ancillaries	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
						F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
				CT2750	Implantable cardiac electrical stimulators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
						F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
				CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
						F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
				CT2985	Active	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		

							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT3001	Implantable metal	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT3002	Implantable synthetic polymer	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT3006	Implantable inorganic material	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT314	Electrically powered	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT321	Long-term surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT449	Generators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT981	Single use	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT983	Surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
7977	2022	All other reportable incidents	35852	Single-chamber implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	17/10/2022	17/10/2022
									A0908	Incorrect, Inadequate or Imprecise Result or Readings	17/10/2022	17/10/2022
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	17/10/2022	17/10/2022
									A0908	Incorrect, Inadequate or Imprecise Result or Readings	17/10/2022	17/10/2022
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	17/10/2022	17/10/2022
									A0908	Incorrect, Inadequate or Imprecise Result or Readings	17/10/2022	17/10/2022
7978	2022	All other reportable incidents	47265	Dual-chamber implantable pacemaker, rate-responsive	CT141	Electrodes/Leads and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	29/09/2022	29/09/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	29/09/2022	29/09/2022
					CT145	Cardiovascular devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	29/09/2022	29/09/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	29/09/2022	29/09/2022
					CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	29/09/2022	29/09/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	29/09/2022	29/09/2022
					CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	29/09/2022	29/09/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	29/09/2022	29/09/2022
					CT233	Surgical	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	29/09/2022	29/09/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	29/09/2022	29/09/2022
					CT2406	Implantable	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	29/09/2022	29/09/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	29/09/2022	29/09/2022
					CT2407	Active implantable	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	29/09/2022	29/09/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	29/09/2022	29/09/2022
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	29/09/2022	29/09/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	29/09/2022	29/09/2022
					CT2727	Active implant devices/ancillaries	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	29/09/2022	29/09/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	29/09/2022	29/09/2022
					CT2750	Implantable cardiac electrical stimulators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	29/09/2022	29/09/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	29/09/2022	29/09/2022
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	29/09/2022	29/09/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	29/09/2022	29/09/2022
					CT2985	Active	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	29/09/2022	29/09/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	29/09/2022	29/09/2022
					CT3001	Implantable metal	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	29/09/2022	29/09/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	29/09/2022	29/09/2022
					CT3002	Implantable synthetic polymer	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	29/09/2022	29/09/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	29/09/2022	29/09/2022
					CT3006	Implantable inorganic material	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	29/09/2022	29/09/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	29/09/2022	29/09/2022
					CT314	Electrically powered	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	29/09/2022	29/09/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	29/09/2022	29/09/2022
					CT321	Long-term surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	29/09/2022	29/09/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	29/09/2022	29/09/2022
					CT449	Generators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	29/09/2022	29/09/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	29/09/2022	29/09/2022
					CT981	Single use	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	29/09/2022	29/09/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	29/09/2022	29/09/2022
					CT983	Surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	29/09/2022	29/09/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	29/09/2022	29/09/2022
7979	2022		47265	Dual-chamber implantable pacemaker, rate-responsive	CT141	Electrodes/Leads and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	27/09/2022	27/09/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	27/09/2022	27/09/2022
					CT145	Cardiovascular devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	27/09/2022	27/09/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	27/09/2022	27/09/2022
					CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	27/09/2022	27/09/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	27/09/2022	27/09/2022
					CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	27/09/2022	27/09/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	27/09/2022	27/09/2022
					CT233	Surgical	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	27/09/2022	27/09/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	27/09/2022	27/09/2022
					CT2406	Implantable	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	27/09/2022	27/09/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	27/09/2022	27/09/2022
					CT2407	Active implantable	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	27/09/2022	27/09/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	27/09/2022	27/09/2022
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	27/09/2022	27/09/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	27/09/2022	27/09/2022
					CT2727	Active implant devices/ancillaries	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	27/09/2022	27/09/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	27/09/2022	27/09/2022
					CT2750	Implantable cardiac electrical stimulators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	27/09/2022	27/09/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	27/09/2022	27/09/2022
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	27/09/2022	27/09/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	27/09/2022	27/09/2022

					CT2985	Active	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	27/09/2022	27/09/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	27/09/2022	27/09/2022
					CT3001	Implantable metal	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	27/09/2022	27/09/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	27/09/2022	27/09/2022
					CT3002	Implantable synthetic polymer	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	27/09/2022	27/09/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	27/09/2022	27/09/2022
					CT3006	Implantable inorganic material	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	27/09/2022	27/09/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	27/09/2022	27/09/2022
					CT314	Electrically powered	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	27/09/2022	27/09/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	27/09/2022	27/09/2022
					CT321	Long-term surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	27/09/2022	27/09/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	27/09/2022	27/09/2022
					CT449	Generators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	27/09/2022	27/09/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	27/09/2022	27/09/2022
					CT981	Single use	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	27/09/2022	27/09/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	27/09/2022	27/09/2022
					CT983	Surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	27/09/2022	27/09/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	27/09/2022	27/09/2022
7980	2022		47265	Dual-chamber implantable pacemaker, rate-responsive	CT141	Electrodes/Leads and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	29/09/2022	29/09/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	29/09/2022	29/09/2022
					CT145	Cardiovascular devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	29/09/2022	29/09/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	29/09/2022	29/09/2022
					CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	29/09/2022	29/09/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	29/09/2022	29/09/2022
					CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	29/09/2022	29/09/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	29/09/2022	29/09/2022
					CT233	Surgical	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	29/09/2022	29/09/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	29/09/2022	29/09/2022
					CT2406	Implantable	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	29/09/2022	29/09/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	29/09/2022	29/09/2022
					CT2407	Active implantable	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	29/09/2022	29/09/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	29/09/2022	29/09/2022
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	29/09/2022	29/09/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	29/09/2022	29/09/2022
					CT2727	Active implant devices/ancillaries	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	29/09/2022	29/09/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	29/09/2022	29/09/2022
					CT2750	Implantable cardiac electrical stimulators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	29/09/2022	29/09/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	29/09/2022	29/09/2022
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	29/09/2022	29/09/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	29/09/2022	29/09/2022
					CT2985	Active	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	29/09/2022	29/09/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	29/09/2022	29/09/2022
					CT3001	Implantable metal	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	29/09/2022	29/09/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	29/09/2022	29/09/2022
					CT3002	Implantable synthetic polymer	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	29/09/2022	29/09/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	29/09/2022	29/09/2022
					CT3006	Implantable inorganic material	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	29/09/2022	29/09/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	29/09/2022	29/09/2022
					CT314	Electrically powered	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	29/09/2022	29/09/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	29/09/2022	29/09/2022
					CT321	Long-term surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	29/09/2022	29/09/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	29/09/2022	29/09/2022
					CT449	Generators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	29/09/2022	29/09/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	29/09/2022	29/09/2022
					CT981	Single use	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	29/09/2022	29/09/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	29/09/2022	29/09/2022
					CT983	Surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	29/09/2022	29/09/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	29/09/2022	29/09/2022
7981	2022		47265	Dual-chamber implantable pacemaker, rate-responsive	CT141	Electrodes/Leads and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	30/09/2022	30/09/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	30/09/2022	30/09/2022
					CT145	Cardiovascular devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	30/09/2022	30/09/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	30/09/2022	30/09/2022
					CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	30/09/2022	30/09/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	30/09/2022	30/09/2022
					CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	30/09/2022	30/09/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	30/09/2022	30/09/2022
					CT233	Surgical	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	30/09/2022	30/09/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	30/09/2022	30/09/2022
					CT2406	Implantable	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	30/09/2022	30/09/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	30/09/2022	30/09/2022
					CT2407	Active implantable	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	30/09/2022	30/09/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	30/09/2022	30/09/2022
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	30/09/2022	30/09/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	30/09/2022	30/09/2022
					CT2727	Active implant devices/ancillaries	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	30/09/2022	30/09/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	30/09/2022	30/09/2022
					CT2750	Implantable cardiac electrical stimulators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	30/09/2022	30/09/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	30/09/2022	30/09/2022
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	30/09/2022	30/09/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	30/09/2022	30/09/2022
					CT2985	Active	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	30/09/2022	30/09/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	30/09/2022	30/09/2022
					CT3001	Implantable metal	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	30/09/2022	30/09/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	30/09/2022	30/09/2022
					CT3002	Implantable synthetic polymer	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	30/09/2022	30/09/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	30/09/2022	30/09/2022
					CT3006	Implantable inorganic material	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	30/09/2022	30/09/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	30/09/2022	30/09/2022

					CT314	Electrically powered	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	30/09/2022	30/09/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	30/09/2022	30/09/2022
					CT321	Long-term surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	30/09/2022	30/09/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	30/09/2022	30/09/2022
					CT449	Generators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	30/09/2022	30/09/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	30/09/2022	30/09/2022
					CT981	Single use	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	30/09/2022	30/09/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	30/09/2022	30/09/2022
					CT983	Surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	30/09/2022	30/09/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	30/09/2022	30/09/2022
7982	2022	All other reportable incidents	47265	Dual-chamber implantable pacemaker, rate-responsive	CT141	Electrodes/Leads and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	27/09/2022	27/09/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	27/09/2022	27/09/2022
					CT145	Cardiovascular devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	27/09/2022	27/09/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	27/09/2022	27/09/2022
					CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	27/09/2022	27/09/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	27/09/2022	27/09/2022
					CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	27/09/2022	27/09/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	27/09/2022	27/09/2022
					CT233	Surgical	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	27/09/2022	27/09/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	27/09/2022	27/09/2022
					CT2406	Implantable	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	27/09/2022	27/09/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	27/09/2022	27/09/2022
					CT2407	Active implantable	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	27/09/2022	27/09/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	27/09/2022	27/09/2022
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	27/09/2022	27/09/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	27/09/2022	27/09/2022
					CT2727	Active implant devices/ancillaries	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	27/09/2022	27/09/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	27/09/2022	27/09/2022
					CT2750	Implantable cardiac electrical stimulators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	27/09/2022	27/09/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	27/09/2022	27/09/2022
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	27/09/2022	27/09/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	27/09/2022	27/09/2022
					CT2985	Active	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	27/09/2022	27/09/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	27/09/2022	27/09/2022
					CT3001	Implantable metal	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	27/09/2022	27/09/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	27/09/2022	27/09/2022
					CT3002	Implantable synthetic polymer	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	27/09/2022	27/09/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	27/09/2022	27/09/2022
					CT3006	Implantable inorganic material	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	27/09/2022	27/09/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	27/09/2022	27/09/2022
					CT314	Electrically powered	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	27/09/2022	27/09/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	27/09/2022	27/09/2022
					CT321	Long-term surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	27/09/2022	27/09/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	27/09/2022	27/09/2022
					CT449	Generators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	27/09/2022	27/09/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	27/09/2022	27/09/2022
					CT981	Single use	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	27/09/2022	27/09/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	27/09/2022	27/09/2022
					CT983	Surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	27/09/2022	27/09/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	27/09/2022	27/09/2022
7983	2022	All other reportable incidents	47265	Dual-chamber implantable pacemaker, rate-responsive	CT141	Electrodes/Leads and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	19/09/2022	19/09/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	19/09/2022	19/09/2022
					CT145	Cardiovascular devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	19/09/2022	19/09/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	19/09/2022	19/09/2022
					CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	19/09/2022	19/09/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	19/09/2022	19/09/2022
					CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	19/09/2022	19/09/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	19/09/2022	19/09/2022
					CT233	Surgical	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	19/09/2022	19/09/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	19/09/2022	19/09/2022
					CT2406	Implantable	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	19/09/2022	19/09/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	19/09/2022	19/09/2022
					CT2407	Active implantable	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	19/09/2022	19/09/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	19/09/2022	19/09/2022
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	19/09/2022	19/09/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	19/09/2022	19/09/2022
					CT2727	Active implant devices/ancillaries	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	19/09/2022	19/09/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	19/09/2022	19/09/2022
					CT2750	Implantable cardiac electrical stimulators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	19/09/2022	19/09/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	19/09/2022	19/09/2022
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	19/09/2022	19/09/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	19/09/2022	19/09/2022
					CT2985	Active	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	19/09/2022	19/09/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	19/09/2022	19/09/2022
					CT3001	Implantable metal	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	19/09/2022	19/09/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	19/09/2022	19/09/2022
					CT3002	Implantable synthetic polymer	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	19/09/2022	19/09/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	19/09/2022	19/09/2022
					CT3006	Implantable inorganic material	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	19/09/2022	19/09/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	19/09/2022	19/09/2022
					CT314	Electrically powered	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	19/09/2022	19/09/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	19/09/2022	19/09/2022
					CT321	Long-term surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	19/09/2022	19/09/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	19/09/2022	19/09/2022
					CT449	Generators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	19/09/2022	19/09/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	19/09/2022	19/09/2022
					CT981	Single use	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	19/09/2022	19/09/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	19/09/2022	19/09/2022

					CT983	Surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	19/09/2022	19/09/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	19/09/2022	19/09/2022
7984	2022		47265	Dual-chamber implantable pacemaker, rate-responsive	CT141	Electrodes/Leads and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	08/10/2022	08/10/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	08/10/2022	08/10/2022
					CT145	Cardiovascular devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	08/10/2022	08/10/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	08/10/2022	08/10/2022
					CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	08/10/2022	08/10/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	08/10/2022	08/10/2022
					CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	08/10/2022	08/10/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	08/10/2022	08/10/2022
					CT233	Surgical	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	08/10/2022	08/10/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	08/10/2022	08/10/2022
					CT2406	Implantable	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	08/10/2022	08/10/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	08/10/2022	08/10/2022
					CT2407	Active implantable	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	08/10/2022	08/10/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	08/10/2022	08/10/2022
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	08/10/2022	08/10/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	08/10/2022	08/10/2022
					CT2727	Active implant devices/ancillaries	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	08/10/2022	08/10/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	08/10/2022	08/10/2022
					CT2750	Implantable cardiac electrical stimulators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	08/10/2022	08/10/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	08/10/2022	08/10/2022
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	08/10/2022	08/10/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	08/10/2022	08/10/2022
					CT2985	Active	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	08/10/2022	08/10/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	08/10/2022	08/10/2022
					CT3001	Implantable metal	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	08/10/2022	08/10/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	08/10/2022	08/10/2022
					CT3002	Implantable synthetic polymer	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	08/10/2022	08/10/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	08/10/2022	08/10/2022
					CT3006	Implantable inorganic material	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	08/10/2022	08/10/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	08/10/2022	08/10/2022
					CT314	Electrically powered	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	08/10/2022	08/10/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	08/10/2022	08/10/2022
					CT321	Long-term surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	08/10/2022	08/10/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	08/10/2022	08/10/2022
					CT449	Generators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	08/10/2022	08/10/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	08/10/2022	08/10/2022
					CT981	Single use	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	08/10/2022	08/10/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	08/10/2022	08/10/2022
					CT983	Surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	08/10/2022	08/10/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	08/10/2022	08/10/2022
7985	2022	All other reportable incidents	47265	Dual-chamber implantable pacemaker, rate-responsive	CT141	Electrodes/Leads and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	03/10/2022	03/10/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	03/10/2022	03/10/2022
					CT145	Cardiovascular devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	03/10/2022	03/10/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	03/10/2022	03/10/2022
					CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	03/10/2022	03/10/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	03/10/2022	03/10/2022
					CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	03/10/2022	03/10/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	03/10/2022	03/10/2022
					CT233	Surgical	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	03/10/2022	03/10/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	03/10/2022	03/10/2022
					CT2406	Implantable	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	03/10/2022	03/10/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	03/10/2022	03/10/2022
					CT2407	Active implantable	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	03/10/2022	03/10/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	03/10/2022	03/10/2022
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	03/10/2022	03/10/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	03/10/2022	03/10/2022
					CT2727	Active implant devices/ancillaries	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	03/10/2022	03/10/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	03/10/2022	03/10/2022
					CT2750	Implantable cardiac electrical stimulators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	03/10/2022	03/10/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	03/10/2022	03/10/2022
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	03/10/2022	03/10/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	03/10/2022	03/10/2022
					CT2985	Active	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	03/10/2022	03/10/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	03/10/2022	03/10/2022
					CT3001	Implantable metal	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	03/10/2022	03/10/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	03/10/2022	03/10/2022
					CT3002	Implantable synthetic polymer	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	03/10/2022	03/10/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	03/10/2022	03/10/2022
					CT3006	Implantable inorganic material	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	03/10/2022	03/10/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	03/10/2022	03/10/2022
					CT314	Electrically powered	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	03/10/2022	03/10/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	03/10/2022	03/10/2022
					CT321	Long-term surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	03/10/2022	03/10/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	03/10/2022	03/10/2022
					CT449	Generators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	03/10/2022	03/10/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	03/10/2022	03/10/2022
					CT981	Single use	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	03/10/2022	03/10/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	03/10/2022	03/10/2022
					CT983	Surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	03/10/2022	03/10/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	03/10/2022	03/10/2022
7986	2022	Unanticipated serious deterioration in state of health	47265	Dual-chamber implantable pacemaker, rate-responsive	CT141	Electrodes/Leads and associated devices	E011903	Syncope/Fainting	A24	No Apparent Device Problem		
							F1203	Life Threatening Illness or Injury	A24	No Apparent Device Problem		
					CT145	Cardiovascular devices	E011903	Syncope/Fainting	A24	No Apparent Device Problem		
							F1203	Life Threatening Illness or Injury	A24	No Apparent Device Problem		
					CT1684	Cardiac pacemakers and associated devices	E011903	Syncope/Fainting	A24	No Apparent Device Problem		
							F1203	Life Threatening Illness or Injury	A24	No Apparent Device Problem		

					CT1685	Cardiac pacemakers	E011903	Syncope/Fainting	A24	No Apparent Device Problem		
							F1203	Life Threatening Illness or Injury	A24	No Apparent Device Problem		
					CT233	Surgical	E011903	Syncope/Fainting	A24	No Apparent Device Problem		
							F1203	Life Threatening Illness or Injury	A24	No Apparent Device Problem		
					CT2406	Implantable	E011903	Syncope/Fainting	A24	No Apparent Device Problem		
							F1203	Life Threatening Illness or Injury	A24	No Apparent Device Problem		
					CT2407	Active implantable	E011903	Syncope/Fainting	A24	No Apparent Device Problem		
							F1203	Life Threatening Illness or Injury	A24	No Apparent Device Problem		
					CT257	Implantable cardiac pacemakers	E011903	Syncope/Fainting	A24	No Apparent Device Problem		
							F1203	Life Threatening Illness or Injury	A24	No Apparent Device Problem		
					CT2727	Active implant devices/ancillaries	E011903	Syncope/Fainting	A24	No Apparent Device Problem		
							F1203	Life Threatening Illness or Injury	A24	No Apparent Device Problem		
					CT2750	Implantable cardiac electrical stimulators and associated devices	E011903	Syncope/Fainting	A24	No Apparent Device Problem		
							F1203	Life Threatening Illness or Injury	A24	No Apparent Device Problem		
					CT2866	Implantable cardiac electrical stimulators	E011903	Syncope/Fainting	A24	No Apparent Device Problem		
							F1203	Life Threatening Illness or Injury	A24	No Apparent Device Problem		
					CT2985	Active	E011903	Syncope/Fainting	A24	No Apparent Device Problem		
							F1203	Life Threatening Illness or Injury	A24	No Apparent Device Problem		
					CT3001	Implantable metal	E011903	Syncope/Fainting	A24	No Apparent Device Problem		
							F1203	Life Threatening Illness or Injury	A24	No Apparent Device Problem		
					CT3002	Implantable synthetic polymer	E011903	Syncope/Fainting	A24	No Apparent Device Problem		
							F1203	Life Threatening Illness or Injury	A24	No Apparent Device Problem		
					CT3006	Implantable inorganic material	E011903	Syncope/Fainting	A24	No Apparent Device Problem		
							F1203	Life Threatening Illness or Injury	A24	No Apparent Device Problem		
					CT314	Electrically powered	E011903	Syncope/Fainting	A24	No Apparent Device Problem		
							F1203	Life Threatening Illness or Injury	A24	No Apparent Device Problem		
					CT321	Long-term surgical invasive	E011903	Syncope/Fainting	A24	No Apparent Device Problem		
							F1203	Life Threatening Illness or Injury	A24	No Apparent Device Problem		
					CT449	Generators and associated devices	E011903	Syncope/Fainting	A24	No Apparent Device Problem		
							F1203	Life Threatening Illness or Injury	A24	No Apparent Device Problem		
					CT981	Single use	E011903	Syncope/Fainting	A24	No Apparent Device Problem		
							F1203	Life Threatening Illness or Injury	A24	No Apparent Device Problem		
					CT983	Surgical invasive	E011903	Syncope/Fainting	A24	No Apparent Device Problem		
							F1203	Life Threatening Illness or Injury	A24	No Apparent Device Problem		
7987	2022	All other reportable incidents	36241	Pacemaker/defibrillator lead	CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A05	Mechanical Problem	26/01/2022	26/01/2022
									A070101	Failure to Capture	26/01/2022	26/01/2022
									A070910	Undersensing	26/01/2022	26/01/2022
							F1908	Prolonged Surgery	A05	Mechanical Problem	26/01/2022	26/01/2022
									A070101	Failure to Capture	26/01/2022	26/01/2022
									A070910	Undersensing	26/01/2022	26/01/2022
7988	2022	All other reportable incidents	35852	Single-chamber implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A072201	High Impedance		
							F26	No Health Consequences or Impact	A072201	High Impedance		
7989	2022	All other reportable incidents	36241	Pacemaker/defibrillator lead	CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A072201	High Impedance		
							F26	No Health Consequences or Impact	A072201	High Impedance		
7990	2022	All other reportable incidents	47265	Dual-chamber implantable pacemaker, rate-responsive	CT141	Electrodes/Leads and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem	31/03/2022	31/03/2022
									A070504	Premature Discharge of Battery	31/03/2022	31/03/2022
							F08	Hospitalization or Prolonged Hospitalization	A0705	Battery Problem	31/03/2022	31/03/2022
									A070504	Premature Discharge of Battery	31/03/2022	31/03/2022
							F1905	Device Revision or Replacement	A0705	Battery Problem	31/03/2022	31/03/2022
									A070504	Premature Discharge of Battery	31/03/2022	31/03/2022
					CT145	Cardiovascular devices	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem	31/03/2022	31/03/2022
									A070504	Premature Discharge of Battery	31/03/2022	31/03/2022
							F08	Hospitalization or Prolonged Hospitalization	A0705	Battery Problem	31/03/2022	31/03/2022
									A070504	Premature Discharge of Battery	31/03/2022	31/03/2022
							F1905	Device Revision or Replacement	A0705	Battery Problem	31/03/2022	31/03/2022
									A070504	Premature Discharge of Battery	31/03/2022	31/03/2022
					CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem	31/03/2022	31/03/2022
									A070504	Premature Discharge of Battery	31/03/2022	31/03/2022
							F08	Hospitalization or Prolonged Hospitalization	A0705	Battery Problem	31/03/2022	31/03/2022
									A070504	Premature Discharge of Battery	31/03/2022	31/03/2022
							F1905	Device Revision or Replacement	A0705	Battery Problem	31/03/2022	31/03/2022
									A070504	Premature Discharge of Battery	31/03/2022	31/03/2022
					CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem	31/03/2022	31/03/2022
									A070504	Premature Discharge of Battery	31/03/2022	31/03/2022
							F08	Hospitalization or Prolonged Hospitalization	A0705	Battery Problem	31/03/2022	31/03/2022
									A070504	Premature Discharge of Battery	31/03/2022	31/03/2022
							F1905	Device Revision or Replacement	A0705	Battery Problem	31/03/2022	31/03/2022
									A070504	Premature Discharge of Battery	31/03/2022	31/03/2022
					CT233	Surgical	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem	31/03/2022	31/03/2022
									A070504	Premature Discharge of Battery	31/03/2022	31/03/2022
							F08	Hospitalization or Prolonged Hospitalization	A0705	Battery Problem	31/03/2022	31/03/2022
									A070504	Premature Discharge of Battery	31/03/2022	31/03/2022
							F1905	Device Revision or Replacement	A0705	Battery Problem	31/03/2022	31/03/2022
									A070504	Premature Discharge of Battery	31/03/2022	31/03/2022
					CT2406	Implantable	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem	31/03/2022	31/03/2022
									A070504	Premature Discharge of Battery	31/03/2022	31/03/2022
							F08	Hospitalization or Prolonged Hospitalization	A0705	Battery Problem	31/03/2022	31/03/2022
									A070504	Premature Discharge of Battery	31/03/2022	31/03/2022
							F1905	Device Revision or Replacement	A0705	Battery Problem	31/03/2022	31/03/2022
									A070504	Premature Discharge of Battery	31/03/2022	31/03/2022
					CT2407	Active implantable	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem	31/03/2022	31/03/2022
									A070504	Premature Discharge of Battery	31/03/2022	31/03/2022
							F08	Hospitalization or Prolonged Hospitalization	A0705	Battery Problem	31/03/2022	31/03/2022
									A070504	Premature Discharge of Battery	31/03/2022	31/03/2022
							F1905	Device Revision or Replacement	A0705	Battery Problem	31/03/2022	31/03/2022
									A070504	Premature Discharge of Battery	31/03/2022	31/03/2022
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem	31/03/2022	31/03/2022
									A070504	Premature Discharge of Battery	31/03/2022	31/03/2022

							F08	Hospitalization or Prolonged Hospitalization	A0705	Battery Problem	31/03/2022	31/03/2022
									A070504	Premature Discharge of Battery	31/03/2022	31/03/2022
							F1905	Device Revision or Replacement	A0705	Battery Problem	31/03/2022	31/03/2022
									A070504	Premature Discharge of Battery	31/03/2022	31/03/2022
					CT2727	Active implant devices/ancillaries	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem	31/03/2022	31/03/2022
									A070504	Premature Discharge of Battery	31/03/2022	31/03/2022
							F08	Hospitalization or Prolonged Hospitalization	A0705	Battery Problem	31/03/2022	31/03/2022
									A070504	Premature Discharge of Battery	31/03/2022	31/03/2022
							F1905	Device Revision or Replacement	A0705	Battery Problem	31/03/2022	31/03/2022
									A070504	Premature Discharge of Battery	31/03/2022	31/03/2022
					CT2750	Implantable cardiac electrical stimulators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem	31/03/2022	31/03/2022
									A070504	Premature Discharge of Battery	31/03/2022	31/03/2022
							F08	Hospitalization or Prolonged Hospitalization	A0705	Battery Problem	31/03/2022	31/03/2022
									A070504	Premature Discharge of Battery	31/03/2022	31/03/2022
							F1905	Device Revision or Replacement	A0705	Battery Problem	31/03/2022	31/03/2022
									A070504	Premature Discharge of Battery	31/03/2022	31/03/2022
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem	31/03/2022	31/03/2022
									A070504	Premature Discharge of Battery	31/03/2022	31/03/2022
							F08	Hospitalization or Prolonged Hospitalization	A0705	Battery Problem	31/03/2022	31/03/2022
									A070504	Premature Discharge of Battery	31/03/2022	31/03/2022
							F1905	Device Revision or Replacement	A0705	Battery Problem	31/03/2022	31/03/2022
									A070504	Premature Discharge of Battery	31/03/2022	31/03/2022
					CT2985	Active	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem	31/03/2022	31/03/2022
									A070504	Premature Discharge of Battery	31/03/2022	31/03/2022
							F08	Hospitalization or Prolonged Hospitalization	A0705	Battery Problem	31/03/2022	31/03/2022
									A070504	Premature Discharge of Battery	31/03/2022	31/03/2022
							F1905	Device Revision or Replacement	A0705	Battery Problem	31/03/2022	31/03/2022
									A070504	Premature Discharge of Battery	31/03/2022	31/03/2022
					CT3001	Implantable metal	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem	31/03/2022	31/03/2022
									A070504	Premature Discharge of Battery	31/03/2022	31/03/2022
							F08	Hospitalization or Prolonged Hospitalization	A0705	Battery Problem	31/03/2022	31/03/2022
									A070504	Premature Discharge of Battery	31/03/2022	31/03/2022
							F1905	Device Revision or Replacement	A0705	Battery Problem	31/03/2022	31/03/2022
									A070504	Premature Discharge of Battery	31/03/2022	31/03/2022
					CT3002	Implantable synthetic polymer	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem	31/03/2022	31/03/2022
									A070504	Premature Discharge of Battery	31/03/2022	31/03/2022
							F08	Hospitalization or Prolonged Hospitalization	A0705	Battery Problem	31/03/2022	31/03/2022
									A070504	Premature Discharge of Battery	31/03/2022	31/03/2022
							F1905	Device Revision or Replacement	A0705	Battery Problem	31/03/2022	31/03/2022
									A070504	Premature Discharge of Battery	31/03/2022	31/03/2022
					CT3006	Implantable inorganic material	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem	31/03/2022	31/03/2022
									A070504	Premature Discharge of Battery	31/03/2022	31/03/2022
							F08	Hospitalization or Prolonged Hospitalization	A0705	Battery Problem	31/03/2022	31/03/2022
									A070504	Premature Discharge of Battery	31/03/2022	31/03/2022
							F1905	Device Revision or Replacement	A0705	Battery Problem	31/03/2022	31/03/2022
									A070504	Premature Discharge of Battery	31/03/2022	31/03/2022
					CT314	Electrically powered	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem	31/03/2022	31/03/2022
									A070504	Premature Discharge of Battery	31/03/2022	31/03/2022
							F08	Hospitalization or Prolonged Hospitalization	A0705	Battery Problem	31/03/2022	31/03/2022
									A070504	Premature Discharge of Battery	31/03/2022	31/03/2022
							F1905	Device Revision or Replacement	A0705	Battery Problem	31/03/2022	31/03/2022
									A070504	Premature Discharge of Battery	31/03/2022	31/03/2022
					CT321	Long-term surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem	31/03/2022	31/03/2022
									A070504	Premature Discharge of Battery	31/03/2022	31/03/2022
							F08	Hospitalization or Prolonged Hospitalization	A0705	Battery Problem	31/03/2022	31/03/2022
									A070504	Premature Discharge of Battery	31/03/2022	31/03/2022
							F1905	Device Revision or Replacement	A0705	Battery Problem	31/03/2022	31/03/2022
									A070504	Premature Discharge of Battery	31/03/2022	31/03/2022
					CT449	Generators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem	31/03/2022	31/03/2022
									A070504	Premature Discharge of Battery	31/03/2022	31/03/2022
							F08	Hospitalization or Prolonged Hospitalization	A0705	Battery Problem	31/03/2022	31/03/2022
									A070504	Premature Discharge of Battery	31/03/2022	31/03/2022
							F1905	Device Revision or Replacement	A0705	Battery Problem	31/03/2022	31/03/2022
									A070504	Premature Discharge of Battery	31/03/2022	31/03/2022
					CT981	Single use	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem	31/03/2022	31/03/2022
									A070504	Premature Discharge of Battery	31/03/2022	31/03/2022
							F08	Hospitalization or Prolonged Hospitalization	A0705	Battery Problem	31/03/2022	31/03/2022
									A070504	Premature Discharge of Battery	31/03/2022	31/03/2022
							F1905	Device Revision or Replacement	A0705	Battery Problem	31/03/2022	31/03/2022
									A070504	Premature Discharge of Battery	31/03/2022	31/03/2022
					CT983	Surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem	31/03/2022	31/03/2022
									A070504	Premature Discharge of Battery	31/03/2022	31/03/2022
							F08	Hospitalization or Prolonged Hospitalization	A0705	Battery Problem	31/03/2022	31/03/2022
									A070504	Premature Discharge of Battery	31/03/2022	31/03/2022
							F1905	Device Revision or Replacement	A0705	Battery Problem	31/03/2022	31/03/2022
									A070504	Premature Discharge of Battery	31/03/2022	31/03/2022
7991	2022	All other reportable incidents	47267	Single-chamber implantable pacemaker, rate-responsive	CT141	Electrodes/Leads and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	10/02/2022	10/02/2022
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	10/02/2022	10/02/2022
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	10/02/2022	10/02/2022
					CT145	Cardiovascular devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	10/02/2022	10/02/2022
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	10/02/2022	10/02/2022
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	10/02/2022	10/02/2022
					CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	10/02/2022	10/02/2022
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	10/02/2022	10/02/2022
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	10/02/2022	10/02/2022
					CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	10/02/2022	10/02/2022
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	10/02/2022	10/02/2022
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	10/02/2022	10/02/2022

					CT233	Surgical	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	10/02/2022	10/02/2022
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	10/02/2022	10/02/2022
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	10/02/2022	10/02/2022
					CT2406	Implantable	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	10/02/2022	10/02/2022
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	10/02/2022	10/02/2022
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	10/02/2022	10/02/2022
					CT2407	Active implantable	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	10/02/2022	10/02/2022
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	10/02/2022	10/02/2022
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	10/02/2022	10/02/2022
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	10/02/2022	10/02/2022
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	10/02/2022	10/02/2022
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	10/02/2022	10/02/2022
					CT2727	Active implant devices/ancillaries	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	10/02/2022	10/02/2022
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	10/02/2022	10/02/2022
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	10/02/2022	10/02/2022
					CT2750	Implantable cardiac electrical stimulators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	10/02/2022	10/02/2022
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	10/02/2022	10/02/2022
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	10/02/2022	10/02/2022
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	10/02/2022	10/02/2022
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	10/02/2022	10/02/2022
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	10/02/2022	10/02/2022
					CT2985	Active	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	10/02/2022	10/02/2022
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	10/02/2022	10/02/2022
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	10/02/2022	10/02/2022
					CT3001	Implantable metal	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	10/02/2022	10/02/2022
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	10/02/2022	10/02/2022
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	10/02/2022	10/02/2022
					CT3002	Implantable synthetic polymer	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	10/02/2022	10/02/2022
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	10/02/2022	10/02/2022
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	10/02/2022	10/02/2022
					CT3006	Implantable inorganic material	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	10/02/2022	10/02/2022
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	10/02/2022	10/02/2022
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	10/02/2022	10/02/2022
					CT314	Electrically powered	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	10/02/2022	10/02/2022
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	10/02/2022	10/02/2022
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	10/02/2022	10/02/2022
					CT321	Long-term surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	10/02/2022	10/02/2022
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	10/02/2022	10/02/2022
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	10/02/2022	10/02/2022
					CT449	Generators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	10/02/2022	10/02/2022
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	10/02/2022	10/02/2022
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	10/02/2022	10/02/2022
					CT981	Single use	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	10/02/2022	10/02/2022
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	10/02/2022	10/02/2022
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	10/02/2022	10/02/2022
					CT983	Surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	10/02/2022	10/02/2022
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	10/02/2022	10/02/2022
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	10/02/2022	10/02/2022
7992	2022	All other reportable incidents	17882	Physiologic-monitoring defibrillation system	CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070803	Failure to Power Up		
							F27	Problem Identified During Non-Clinical Procedure	A070803	Failure to Power Up		
7993	2022	All other reportable incidents	17882	Physiologic-monitoring defibrillation system	CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070803	Failure to Power Up		
							F27	Problem Identified During Non-Clinical Procedure	A070803	Failure to Power Up		
7994	2022	All other reportable incidents	47265	Dual-chamber implantable pacemaker, rate-responsive	CT141	Electrodes/Leads and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A071102	Failure to Interrogate	07/10/2022	07/10/2022
									A071204	No Pacing	07/10/2022	07/10/2022
							F1901	Additional Surgery	A071102	Failure to Interrogate	07/10/2022	07/10/2022
									A071204	No Pacing	07/10/2022	07/10/2022
					CT145	Cardiovascular devices	E2403	No Clinical Signs, Symptoms or Conditions	A071102	Failure to Interrogate	07/10/2022	07/10/2022
									A071204	No Pacing	07/10/2022	07/10/2022
							F1901	Additional Surgery	A071102	Failure to Interrogate	07/10/2022	07/10/2022
									A071204	No Pacing	07/10/2022	07/10/2022
					CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A071102	Failure to Interrogate	07/10/2022	07/10/2022
									A071204	No Pacing	07/10/2022	07/10/2022
							F1901	Additional Surgery	A071102	Failure to Interrogate	07/10/2022	07/10/2022
									A071204	No Pacing	07/10/2022	07/10/2022
					CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A071102	Failure to Interrogate	07/10/2022	07/10/2022
									A071204	No Pacing	07/10/2022	07/10/2022
							F1901	Additional Surgery	A071102	Failure to Interrogate	07/10/2022	07/10/2022
									A071204	No Pacing	07/10/2022	07/10/2022
					CT233	Surgical	E2403	No Clinical Signs, Symptoms or Conditions	A071102	Failure to Interrogate	07/10/2022	07/10/2022
									A071204	No Pacing	07/10/2022	07/10/2022
							F1901	Additional Surgery	A071102	Failure to Interrogate	07/10/2022	07/10/2022
									A071204	No Pacing	07/10/2022	07/10/2022
					CT2406	Implantable	E2403	No Clinical Signs, Symptoms or Conditions	A071102	Failure to Interrogate	07/10/2022	07/10/2022
									A071204	No Pacing	07/10/2022	07/10/2022
							F1901	Additional Surgery	A071102	Failure to Interrogate	07/10/2022	07/10/2022
									A071204	No Pacing	07/10/2022	07/10/2022
					CT2407	Active implantable	E2403	No Clinical Signs, Symptoms or Conditions	A071102	Failure to Interrogate	07/10/2022	07/10/2022
									A071204	No Pacing	07/10/2022	07/10/2022
							F1901	Additional Surgery	A071102	Failure to Interrogate	07/10/2022	07/10/2022
									A071204	No Pacing	07/10/2022	07/10/2022
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A071102	Failure to Interrogate	07/10/2022	07/10/2022
									A071204	No Pacing	07/10/2022	07/10/2022
							F1901	Additional Surgery	A071102	Failure to Interrogate	07/10/2022	07/10/2022
									A071204	No Pacing	07/10/2022	07/10/2022
					CT2727	Active implant devices/ancillaries	E2403	No Clinical Signs, Symptoms or Conditions	A071102	Failure to Interrogate	07/10/2022	07/10/2022
									A071204	No Pacing	07/10/2022	07/10/2022
							F1901	Additional Surgery	A071102	Failure to Interrogate	07/10/2022	07/10/2022
									A071204	No Pacing	07/10/2022	07/10/2022

					CT2750	Implantable cardiac electrical stimulators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A071102	Failure to Interrogate	07/10/2022	07/10/2022
									A071204	No Pacing	07/10/2022	07/10/2022
							F1901	Additional Surgery	A071102	Failure to Interrogate	07/10/2022	07/10/2022
									A071204	No Pacing	07/10/2022	07/10/2022
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A071102	Failure to Interrogate	07/10/2022	07/10/2022
									A071204	No Pacing	07/10/2022	07/10/2022
							F1901	Additional Surgery	A071102	Failure to Interrogate	07/10/2022	07/10/2022
									A071204	No Pacing	07/10/2022	07/10/2022
					CT2985	Active	E2403	No Clinical Signs, Symptoms or Conditions	A071102	Failure to Interrogate	07/10/2022	07/10/2022
									A071204	No Pacing	07/10/2022	07/10/2022
							F1901	Additional Surgery	A071102	Failure to Interrogate	07/10/2022	07/10/2022
									A071204	No Pacing	07/10/2022	07/10/2022
					CT3001	Implantable metal	E2403	No Clinical Signs, Symptoms or Conditions	A071102	Failure to Interrogate	07/10/2022	07/10/2022
									A071204	No Pacing	07/10/2022	07/10/2022
							F1901	Additional Surgery	A071102	Failure to Interrogate	07/10/2022	07/10/2022
									A071204	No Pacing	07/10/2022	07/10/2022
					CT3002	Implantable synthetic polymer	E2403	No Clinical Signs, Symptoms or Conditions	A071102	Failure to Interrogate	07/10/2022	07/10/2022
									A071204	No Pacing	07/10/2022	07/10/2022
							F1901	Additional Surgery	A071102	Failure to Interrogate	07/10/2022	07/10/2022
									A071204	No Pacing	07/10/2022	07/10/2022
					CT3006	Implantable inorganic material	E2403	No Clinical Signs, Symptoms or Conditions	A071102	Failure to Interrogate	07/10/2022	07/10/2022
									A071204	No Pacing	07/10/2022	07/10/2022
							F1901	Additional Surgery	A071102	Failure to Interrogate	07/10/2022	07/10/2022
									A071204	No Pacing	07/10/2022	07/10/2022
					CT314	Electrically powered	E2403	No Clinical Signs, Symptoms or Conditions	A071102	Failure to Interrogate	07/10/2022	07/10/2022
									A071204	No Pacing	07/10/2022	07/10/2022
							F1901	Additional Surgery	A071102	Failure to Interrogate	07/10/2022	07/10/2022
									A071204	No Pacing	07/10/2022	07/10/2022
					CT321	Long-term surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A071102	Failure to Interrogate	07/10/2022	07/10/2022
									A071204	No Pacing	07/10/2022	07/10/2022
							F1901	Additional Surgery	A071102	Failure to Interrogate	07/10/2022	07/10/2022
									A071204	No Pacing	07/10/2022	07/10/2022
					CT449	Generators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A071102	Failure to Interrogate	07/10/2022	07/10/2022
									A071204	No Pacing	07/10/2022	07/10/2022
							F1901	Additional Surgery	A071102	Failure to Interrogate	07/10/2022	07/10/2022
									A071204	No Pacing	07/10/2022	07/10/2022
					CT981	Single use	E2403	No Clinical Signs, Symptoms or Conditions	A071102	Failure to Interrogate	07/10/2022	07/10/2022
									A071204	No Pacing	07/10/2022	07/10/2022
							F1901	Additional Surgery	A071102	Failure to Interrogate	07/10/2022	07/10/2022
									A071204	No Pacing	07/10/2022	07/10/2022
					CT983	Surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A071102	Failure to Interrogate	07/10/2022	07/10/2022
									A071204	No Pacing	07/10/2022	07/10/2022
							F1901	Additional Surgery	A071102	Failure to Interrogate	07/10/2022	07/10/2022
									A071204	No Pacing	07/10/2022	07/10/2022
7995	2022	All other reportable incidents	47267	Single-chamber implantable pacemaker, rate-responsive	CT141	Electrodes/Leads and associated devices	E011903	Syncope/Fainting	A070101	Failure to Capture	03/10/2022	03/10/2022
							E2403	No Clinical Signs, Symptoms or Conditions	A070101	Failure to Capture	03/10/2022	03/10/2022
							F08	Hospitalization or Prolonged Hospitalization	A070101	Failure to Capture	03/10/2022	03/10/2022
							F1203	Life Threatening Illness or Injury	A070101	Failure to Capture	03/10/2022	03/10/2022
							F1903	Device Explantation	A070101	Failure to Capture	03/10/2022	03/10/2022
					CT145	Cardiovascular devices	E011903	Syncope/Fainting	A070101	Failure to Capture	03/10/2022	03/10/2022
							E2403	No Clinical Signs, Symptoms or Conditions	A070101	Failure to Capture	03/10/2022	03/10/2022
							F08	Hospitalization or Prolonged Hospitalization	A070101	Failure to Capture	03/10/2022	03/10/2022
							F1203	Life Threatening Illness or Injury	A070101	Failure to Capture	03/10/2022	03/10/2022
							F1903	Device Explantation	A070101	Failure to Capture	03/10/2022	03/10/2022
					CT1684	Cardiac pacemakers and associated devices	E011903	Syncope/Fainting	A070101	Failure to Capture	03/10/2022	03/10/2022
							E2403	No Clinical Signs, Symptoms or Conditions	A070101	Failure to Capture	03/10/2022	03/10/2022
							F08	Hospitalization or Prolonged Hospitalization	A070101	Failure to Capture	03/10/2022	03/10/2022
							F1203	Life Threatening Illness or Injury	A070101	Failure to Capture	03/10/2022	03/10/2022
							F1903	Device Explantation	A070101	Failure to Capture	03/10/2022	03/10/2022
					CT1685	Cardiac pacemakers	E011903	Syncope/Fainting	A070101	Failure to Capture	03/10/2022	03/10/2022
							E2403	No Clinical Signs, Symptoms or Conditions	A070101	Failure to Capture	03/10/2022	03/10/2022
							F08	Hospitalization or Prolonged Hospitalization	A070101	Failure to Capture	03/10/2022	03/10/2022
							F1203	Life Threatening Illness or Injury	A070101	Failure to Capture	03/10/2022	03/10/2022
							F1903	Device Explantation	A070101	Failure to Capture	03/10/2022	03/10/2022
					CT233	Surgical	E011903	Syncope/Fainting	A070101	Failure to Capture	03/10/2022	03/10/2022
							E2403	No Clinical Signs, Symptoms or Conditions	A070101	Failure to Capture	03/10/2022	03/10/2022
							F08	Hospitalization or Prolonged Hospitalization	A070101	Failure to Capture	03/10/2022	03/10/2022
							F1203	Life Threatening Illness or Injury	A070101	Failure to Capture	03/10/2022	03/10/2022
							F1903	Device Explantation	A070101	Failure to Capture	03/10/2022	03/10/2022
					CT2406	Implantable	E011903	Syncope/Fainting	A070101	Failure to Capture	03/10/2022	03/10/2022
							E2403	No Clinical Signs, Symptoms or Conditions	A070101	Failure to Capture	03/10/2022	03/10/2022
							F08	Hospitalization or Prolonged Hospitalization	A070101	Failure to Capture	03/10/2022	03/10/2022
							F1203	Life Threatening Illness or Injury	A070101	Failure to Capture	03/10/2022	03/10/2022
							F1903	Device Explantation	A070101	Failure to Capture	03/10/2022	03/10/2022
					CT2407	Active implantable	E011903	Syncope/Fainting	A070101	Failure to Capture	03/10/2022	03/10/2022
							E2403	No Clinical Signs, Symptoms or Conditions	A070101	Failure to Capture	03/10/2022	03/10/2022
							F08	Hospitalization or Prolonged Hospitalization	A070101	Failure to Capture	03/10/2022	03/10/2022
							F1203	Life Threatening Illness or Injury	A070101	Failure to Capture	03/10/2022	03/10/2022
							F1903	Device Explantation	A070101	Failure to Capture	03/10/2022	03/10/2022
					CT257	Implantable cardiac pacemakers	E011903	Syncope/Fainting	A070101	Failure to Capture	03/10/2022	03/10/2022
							E2403	No Clinical Signs, Symptoms or Conditions	A070101	Failure to Capture	03/10/2022	03/10/2022
							F08	Hospitalization or Prolonged Hospitalization	A070101	Failure to Capture	03/10/2022	03/10/2022
							F1203	Life Threatening Illness or Injury	A070101	Failure to Capture	03/10/2022	03/10/2022
							F1903	Device Explantation	A070101	Failure to Capture	03/10/2022	03/10/2022
					CT2727	Active implant devices/ancillaries	E011903	Syncope/Fainting	A070101	Failure to Capture	03/10/2022	03/10/2022
							E2403	No Clinical Signs, Symptoms or Conditions	A070101	Failure to Capture	03/10/2022	03/10/2022
							F08	Hospitalization or Prolonged Hospitalization	A070101	Failure to Capture	03/10/2022	03/10/2022
							F1203	Life Threatening Illness or Injury	A070101	Failure to Capture	03/10/2022	03/10/2022

							F1903	Device Explantation	A070101	Failure to Capture	03/10/2022	03/10/2022
					CT2750	Implantable cardiac electrical stimulators and associated devices	E011903	Syncope/Fainting	A070101	Failure to Capture	03/10/2022	03/10/2022
							E2403	No Clinical Signs, Symptoms or Conditions	A070101	Failure to Capture	03/10/2022	03/10/2022
							F08	Hospitalization or Prolonged Hospitalization	A070101	Failure to Capture	03/10/2022	03/10/2022
							F1203	Life Threatening Illness or Injury	A070101	Failure to Capture	03/10/2022	03/10/2022
							F1903	Device Explantation	A070101	Failure to Capture	03/10/2022	03/10/2022
					CT2866	Implantable cardiac electrical stimulators	E011903	Syncope/Fainting	A070101	Failure to Capture	03/10/2022	03/10/2022
							E2403	No Clinical Signs, Symptoms or Conditions	A070101	Failure to Capture	03/10/2022	03/10/2022
							F08	Hospitalization or Prolonged Hospitalization	A070101	Failure to Capture	03/10/2022	03/10/2022
							F1203	Life Threatening Illness or Injury	A070101	Failure to Capture	03/10/2022	03/10/2022
							F1903	Device Explantation	A070101	Failure to Capture	03/10/2022	03/10/2022
					CT2985	Active	E011903	Syncope/Fainting	A070101	Failure to Capture	03/10/2022	03/10/2022
							E2403	No Clinical Signs, Symptoms or Conditions	A070101	Failure to Capture	03/10/2022	03/10/2022
							F08	Hospitalization or Prolonged Hospitalization	A070101	Failure to Capture	03/10/2022	03/10/2022
							F1203	Life Threatening Illness or Injury	A070101	Failure to Capture	03/10/2022	03/10/2022
							F1903	Device Explantation	A070101	Failure to Capture	03/10/2022	03/10/2022
					CT3001	Implantable metal	E011903	Syncope/Fainting	A070101	Failure to Capture	03/10/2022	03/10/2022
							E2403	No Clinical Signs, Symptoms or Conditions	A070101	Failure to Capture	03/10/2022	03/10/2022
							F08	Hospitalization or Prolonged Hospitalization	A070101	Failure to Capture	03/10/2022	03/10/2022
							F1203	Life Threatening Illness or Injury	A070101	Failure to Capture	03/10/2022	03/10/2022
							F1903	Device Explantation	A070101	Failure to Capture	03/10/2022	03/10/2022
					CT3002	Implantable synthetic polymer	E011903	Syncope/Fainting	A070101	Failure to Capture	03/10/2022	03/10/2022
							E2403	No Clinical Signs, Symptoms or Conditions	A070101	Failure to Capture	03/10/2022	03/10/2022
							F08	Hospitalization or Prolonged Hospitalization	A070101	Failure to Capture	03/10/2022	03/10/2022
							F1203	Life Threatening Illness or Injury	A070101	Failure to Capture	03/10/2022	03/10/2022
							F1903	Device Explantation	A070101	Failure to Capture	03/10/2022	03/10/2022
					CT3006	Implantable inorganic material	E011903	Syncope/Fainting	A070101	Failure to Capture	03/10/2022	03/10/2022
							E2403	No Clinical Signs, Symptoms or Conditions	A070101	Failure to Capture	03/10/2022	03/10/2022
							F08	Hospitalization or Prolonged Hospitalization	A070101	Failure to Capture	03/10/2022	03/10/2022
							F1203	Life Threatening Illness or Injury	A070101	Failure to Capture	03/10/2022	03/10/2022
							F1903	Device Explantation	A070101	Failure to Capture	03/10/2022	03/10/2022
					CT314	Electrically powered	E011903	Syncope/Fainting	A070101	Failure to Capture	03/10/2022	03/10/2022
							E2403	No Clinical Signs, Symptoms or Conditions	A070101	Failure to Capture	03/10/2022	03/10/2022
							F08	Hospitalization or Prolonged Hospitalization	A070101	Failure to Capture	03/10/2022	03/10/2022
							F1203	Life Threatening Illness or Injury	A070101	Failure to Capture	03/10/2022	03/10/2022
							F1903	Device Explantation	A070101	Failure to Capture	03/10/2022	03/10/2022
					CT321	Long-term surgical invasive	E011903	Syncope/Fainting	A070101	Failure to Capture	03/10/2022	03/10/2022
							E2403	No Clinical Signs, Symptoms or Conditions	A070101	Failure to Capture	03/10/2022	03/10/2022
							F08	Hospitalization or Prolonged Hospitalization	A070101	Failure to Capture	03/10/2022	03/10/2022
							F1203	Life Threatening Illness or Injury	A070101	Failure to Capture	03/10/2022	03/10/2022
							F1903	Device Explantation	A070101	Failure to Capture	03/10/2022	03/10/2022
					CT449	Generators and associated devices	E011903	Syncope/Fainting	A070101	Failure to Capture	03/10/2022	03/10/2022
							E2403	No Clinical Signs, Symptoms or Conditions	A070101	Failure to Capture	03/10/2022	03/10/2022
							F08	Hospitalization or Prolonged Hospitalization	A070101	Failure to Capture	03/10/2022	03/10/2022
							F1203	Life Threatening Illness or Injury	A070101	Failure to Capture	03/10/2022	03/10/2022
							F1903	Device Explantation	A070101	Failure to Capture	03/10/2022	03/10/2022
					CT981	Single use	E011903	Syncope/Fainting	A070101	Failure to Capture	03/10/2022	03/10/2022
							E2403	No Clinical Signs, Symptoms or Conditions	A070101	Failure to Capture	03/10/2022	03/10/2022
							F08	Hospitalization or Prolonged Hospitalization	A070101	Failure to Capture	03/10/2022	03/10/2022
							F1203	Life Threatening Illness or Injury	A070101	Failure to Capture	03/10/2022	03/10/2022
							F1903	Device Explantation	A070101	Failure to Capture	03/10/2022	03/10/2022
					CT983	Surgical invasive	E011903	Syncope/Fainting	A070101	Failure to Capture	03/10/2022	03/10/2022
							E2403	No Clinical Signs, Symptoms or Conditions	A070101	Failure to Capture	03/10/2022	03/10/2022
							F08	Hospitalization or Prolonged Hospitalization	A070101	Failure to Capture	03/10/2022	03/10/2022
							F1203	Life Threatening Illness or Injury	A070101	Failure to Capture	03/10/2022	03/10/2022
							F1903	Device Explantation	A070101	Failure to Capture	03/10/2022	03/10/2022
7996	2022		47265	Dual-chamber implantable pacemaker, rate-responsive	CT141	Electrodes/Leads and associated devices	E0601	Arrhythmia	A24	No Apparent Device Problem		
							F02	Death	A24	No Apparent Device Problem		
					CT145	Cardiovascular devices	E0601	Arrhythmia	A24	No Apparent Device Problem		
							F02	Death	A24	No Apparent Device Problem		
					CT1684	Cardiac pacemakers and associated devices	E0601	Arrhythmia	A24	No Apparent Device Problem		
							F02	Death	A24	No Apparent Device Problem		
					CT1685	Cardiac pacemakers	E0601	Arrhythmia	A24	No Apparent Device Problem		
							F02	Death	A24	No Apparent Device Problem		
					CT233	Surgical	E0601	Arrhythmia	A24	No Apparent Device Problem		
							F02	Death	A24	No Apparent Device Problem		
					CT2406	Implantable	E0601	Arrhythmia	A24	No Apparent Device Problem		
							F02	Death	A24	No Apparent Device Problem		
					CT2407	Active implantable	E0601	Arrhythmia	A24	No Apparent Device Problem		
							F02	Death	A24	No Apparent Device Problem		
					CT257	Implantable cardiac pacemakers	E0601	Arrhythmia	A24	No Apparent Device Problem		
							F02	Death	A24	No Apparent Device Problem		
					CT2727	Active implant devices/ancillaries	E0601	Arrhythmia	A24	No Apparent Device Problem		
							F02	Death	A24	No Apparent Device Problem		
					CT2750	Implantable cardiac electrical stimulators and associated devices	E0601	Arrhythmia	A24	No Apparent Device Problem		
							F02	Death	A24	No Apparent Device Problem		
					CT2866	Implantable cardiac electrical stimulators	E0601	Arrhythmia	A24	No Apparent Device Problem		
							F02	Death	A24	No Apparent Device Problem		
					CT2985	Active	E0601	Arrhythmia	A24	No Apparent Device Problem		
							F02	Death	A24	No Apparent Device Problem		
					CT3001	Implantable metal	E0601	Arrhythmia	A24	No Apparent Device Problem		
							F02	Death	A24	No Apparent Device Problem		
					CT3002	Implantable synthetic polymer	E0601	Arrhythmia	A24	No Apparent Device Problem		
							F02	Death	A24	No Apparent Device Problem		
					CT3006	Implantable inorganic material	E0601	Arrhythmia	A24	No Apparent Device Problem		
							F02	Death	A24	No Apparent Device Problem		
					CT314	Electrically powered	E0601	Arrhythmia	A24	No Apparent Device Problem		
							F02	Death	A24	No Apparent Device Problem		

					CT321	Long-term surgical invasive	E0601	Arrhythmia	A24	No Apparent Device Problem		
							F02	Death	A24	No Apparent Device Problem		
					CT449	Generators and associated devices	E0601	Arrhythmia	A24	No Apparent Device Problem		
							F02	Death	A24	No Apparent Device Problem		
					CT981	Single use	E0601	Arrhythmia	A24	No Apparent Device Problem		
							F02	Death	A24	No Apparent Device Problem		
					CT983	Surgical invasive	E0601	Arrhythmia	A24	No Apparent Device Problem		
							F02	Death	A24	No Apparent Device Problem		
7997	2022		35223	Endocardial/interventricular septal pacing lead	CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070910	Undersensing		
									A072201	High Impedance		
							F02	Death	A070910	Undersensing		
									A072201	High Impedance		
7998	2022		47265	Dual-chamber implantable pacemaker, rate-responsive	CT141	Electrodes/Leads and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A071206	Pacing Asynchronously		
							F02	Death	A071206	Pacing Asynchronously		
					CT145	Cardiovascular devices	E2403	No Clinical Signs, Symptoms or Conditions	A071206	Pacing Asynchronously		
							F02	Death	A071206	Pacing Asynchronously		
					CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A071206	Pacing Asynchronously		
							F02	Death	A071206	Pacing Asynchronously		
					CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A071206	Pacing Asynchronously		
							F02	Death	A071206	Pacing Asynchronously		
					CT233	Surgical	E2403	No Clinical Signs, Symptoms or Conditions	A071206	Pacing Asynchronously		
							F02	Death	A071206	Pacing Asynchronously		
					CT2406	Implantable	E2403	No Clinical Signs, Symptoms or Conditions	A071206	Pacing Asynchronously		
							F02	Death	A071206	Pacing Asynchronously		
					CT2407	Active implantable	E2403	No Clinical Signs, Symptoms or Conditions	A071206	Pacing Asynchronously		
							F02	Death	A071206	Pacing Asynchronously		
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A071206	Pacing Asynchronously		
							F02	Death	A071206	Pacing Asynchronously		
					CT2727	Active implant devices/ancillaries	E2403	No Clinical Signs, Symptoms or Conditions	A071206	Pacing Asynchronously		
							F02	Death	A071206	Pacing Asynchronously		
					CT2750	Implantable cardiac electrical stimulators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A071206	Pacing Asynchronously		
							F02	Death	A071206	Pacing Asynchronously		
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A071206	Pacing Asynchronously		
							F02	Death	A071206	Pacing Asynchronously		
					CT2985	Active	E2403	No Clinical Signs, Symptoms or Conditions	A071206	Pacing Asynchronously		
							F02	Death	A071206	Pacing Asynchronously		
					CT3001	Implantable metal	E2403	No Clinical Signs, Symptoms or Conditions	A071206	Pacing Asynchronously		
							F02	Death	A071206	Pacing Asynchronously		
					CT3002	Implantable synthetic polymer	E2403	No Clinical Signs, Symptoms or Conditions	A071206	Pacing Asynchronously		
							F02	Death	A071206	Pacing Asynchronously		
					CT3006	Implantable inorganic material	E2403	No Clinical Signs, Symptoms or Conditions	A071206	Pacing Asynchronously		
							F02	Death	A071206	Pacing Asynchronously		
					CT314	Electrically powered	E2403	No Clinical Signs, Symptoms or Conditions	A071206	Pacing Asynchronously		
							F02	Death	A071206	Pacing Asynchronously		
					CT321	Long-term surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A071206	Pacing Asynchronously		
							F02	Death	A071206	Pacing Asynchronously		
					CT449	Generators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A071206	Pacing Asynchronously		
							F02	Death	A071206	Pacing Asynchronously		
					CT981	Single use	E2403	No Clinical Signs, Symptoms or Conditions	A071206	Pacing Asynchronously		
							F02	Death	A071206	Pacing Asynchronously		
					CT983	Surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A071206	Pacing Asynchronously		
							F02	Death	A071206	Pacing Asynchronously		
7999	2022	All other reportable incidents	17882	Physiologic-monitoring defibrillation system	CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070803	Failure to Power Up		
							F27	Problem Identified During Non-Clinical Procedure	A070803	Failure to Power Up		
8000	2022	All other reportable incidents	17882	Physiologic-monitoring defibrillation system	CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A04	Material Integrity Problem		
							F27	Problem Identified During Non-Clinical Procedure	A04	Material Integrity Problem		
8001	2022	All other reportable incidents	37265	Dual-chamber implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
8002	2022	All other reportable incidents	37265	Dual-chamber implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	16/02/2022	16/02/2022
							F1901	Additional Surgery	A070504	Premature Discharge of Battery	16/02/2022	16/02/2022
8003	2022	All other reportable incidents	47270	Cardiac resynchronization therapy implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A072201	High Impedance	09/02/2022	09/02/2022
							F1901	Additional Surgery	A072201	High Impedance	09/02/2022	09/02/2022
8004	2022	All other reportable incidents	47267	Single-chamber implantable pacemaker, rate-responsive	CT141	Electrodes/Leads and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A0701	Capturing Problem		
									A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A0701	Capturing Problem		
									A070504	Premature Discharge of Battery		
					CT145	Cardiovascular devices	E2403	No Clinical Signs, Symptoms or Conditions	A0701	Capturing Problem		
									A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A0701	Capturing Problem		
									A070504	Premature Discharge of Battery		
					CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A0701	Capturing Problem		
									A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A0701	Capturing Problem		
									A070504	Premature Discharge of Battery		
					CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A0701	Capturing Problem		
									A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A0701	Capturing Problem		
									A070504	Premature Discharge of Battery		
					CT233	Surgical	E2403	No Clinical Signs, Symptoms or Conditions	A0701	Capturing Problem		
									A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A0701	Capturing Problem		
									A070504	Premature Discharge of Battery		
					CT2406	Implantable	E2403	No Clinical Signs, Symptoms or Conditions	A0701	Capturing Problem		
									A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A0701	Capturing Problem		
									A070504	Premature Discharge of Battery		
					CT2407	Active implantable	E2403	No Clinical Signs, Symptoms or Conditions	A0701	Capturing Problem		
									A070504	Premature Discharge of Battery		

							F26	No Health Consequences or Impact	A0701	Capturing Problem		
									A070504	Premature Discharge of Battery		
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A0701	Capturing Problem		
									A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A0701	Capturing Problem		
									A070504	Premature Discharge of Battery		
					CT2727	Active implant devices/ancillaries	E2403	No Clinical Signs, Symptoms or Conditions	A0701	Capturing Problem		
									A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A0701	Capturing Problem		
									A070504	Premature Discharge of Battery		
					CT2750	Implantable cardiac electrical stimulators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A0701	Capturing Problem		
									A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A0701	Capturing Problem		
									A070504	Premature Discharge of Battery		
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A0701	Capturing Problem		
									A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A0701	Capturing Problem		
									A070504	Premature Discharge of Battery		
					CT2985	Active	E2403	No Clinical Signs, Symptoms or Conditions	A0701	Capturing Problem		
									A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A0701	Capturing Problem		
									A070504	Premature Discharge of Battery		
					CT3001	Implantable metal	E2403	No Clinical Signs, Symptoms or Conditions	A0701	Capturing Problem		
									A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A0701	Capturing Problem		
									A070504	Premature Discharge of Battery		
					CT3002	Implantable synthetic polymer	E2403	No Clinical Signs, Symptoms or Conditions	A0701	Capturing Problem		
									A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A0701	Capturing Problem		
									A070504	Premature Discharge of Battery		
					CT3006	Implantable inorganic material	E2403	No Clinical Signs, Symptoms or Conditions	A0701	Capturing Problem		
									A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A0701	Capturing Problem		
									A070504	Premature Discharge of Battery		
					CT314	Electrically powered	E2403	No Clinical Signs, Symptoms or Conditions	A0701	Capturing Problem		
									A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A0701	Capturing Problem		
									A070504	Premature Discharge of Battery		
					CT321	Long-term surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A0701	Capturing Problem		
									A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A0701	Capturing Problem		
									A070504	Premature Discharge of Battery		
					CT449	Generators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A0701	Capturing Problem		
									A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A0701	Capturing Problem		
									A070504	Premature Discharge of Battery		
					CT981	Single use	E2403	No Clinical Signs, Symptoms or Conditions	A0701	Capturing Problem		
									A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A0701	Capturing Problem		
									A070504	Premature Discharge of Battery		
					CT983	Surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A0701	Capturing Problem		
									A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A0701	Capturing Problem		
									A070504	Premature Discharge of Battery		
8005	2022	All other reportable incidents	35223	Endocardial/interventricular septal pacing lead	CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070102	High Capture Threshold		
									A070103	Intermittent Capture		
									A0722	Impedance Problem		
							F23	Unexpected Medical Intervention	A070102	High Capture Threshold		
									A070103	Intermittent Capture		
									A0722	Impedance Problem		
8006	2022	All other reportable incidents	47270	Cardiac resynchronization therapy implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A0712	Pacing Problem		
							F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A0712	Pacing Problem		
8007	2022	All other reportable incidents	35223	Endocardial/interventricular septal pacing lead	CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A072202	Low Impedance		
							F26	No Health Consequences or Impact	A072202	Low Impedance		
8008	2022	Unanticipated serious deterioration in state of health	47265	Dual-chamber implantable pacemaker, rate-responsive	CT141	Electrodes/Leads and associated devices	E0112	Dizziness	A070101	Failure to Capture		
									A070103	Intermittent Capture		
									A0712	Pacing Problem		
							E011903	Syncope/Fainting	A070101	Failure to Capture		
									A070103	Intermittent Capture		
									A0712	Pacing Problem		
							E060101	Asystole	A070101	Failure to Capture		
									A070103	Intermittent Capture		
									A0712	Pacing Problem		
							F08	Hospitalization or Prolonged Hospitalization	A070101	Failure to Capture		
									A070103	Intermittent Capture		
									A0712	Pacing Problem		
							F1203	Life Threatening Illness or Injury	A070101	Failure to Capture		
									A070103	Intermittent Capture		
									A0712	Pacing Problem		
							F23	Unexpected Medical Intervention	A070101	Failure to Capture		
									A070103	Intermittent Capture		
									A0712	Pacing Problem		
					CT145	Cardiovascular devices	E0112	Dizziness	A070101	Failure to Capture		
									A070103	Intermittent Capture		
									A0712	Pacing Problem		
							E011903	Syncope/Fainting	A070101	Failure to Capture		
									A070103	Intermittent Capture		
									A0712	Pacing Problem		

							E060101	Asystole	A070101	Failure to Capture		
									A070103	Intermittent Capture		
									A0712	Pacing Problem		
							F08	Hospitalization or Prolonged Hospitalization	A070101	Failure to Capture		
									A070103	Intermittent Capture		
									A0712	Pacing Problem		
							F1203	Life Threatening Illness or Injury	A070101	Failure to Capture		
									A070103	Intermittent Capture		
									A0712	Pacing Problem		
							F23	Unexpected Medical Intervention	A070101	Failure to Capture		
									A070103	Intermittent Capture		
									A0712	Pacing Problem		
						CT1684	Cardiac pacemakers and associated devices	E0112	Dizziness	A070101	Failure to Capture		
										A070103	Intermittent Capture		
										A0712	Pacing Problem		
								E011903	Syncope/Fainting	A070101	Failure to Capture		
										A070103	Intermittent Capture		
										A0712	Pacing Problem		
								E060101	Asystole	A070101	Failure to Capture		
										A070103	Intermittent Capture		
										A0712	Pacing Problem		
								F08	Hospitalization or Prolonged Hospitalization	A070101	Failure to Capture		
										A070103	Intermittent Capture		
										A0712	Pacing Problem		
								F1203	Life Threatening Illness or Injury	A070101	Failure to Capture		
										A070103	Intermittent Capture		
										A0712	Pacing Problem		
								F23	Unexpected Medical Intervention	A070101	Failure to Capture		
										A070103	Intermittent Capture		
										A0712	Pacing Problem		
						CT1685	Cardiac pacemakers	E0112	Dizziness	A070101	Failure to Capture		
										A070103	Intermittent Capture		
										A0712	Pacing Problem		
								E011903	Syncope/Fainting	A070101	Failure to Capture		
										A070103	Intermittent Capture		
										A0712	Pacing Problem		
								E060101	Asystole	A070101	Failure to Capture		
										A070103	Intermittent Capture		
										A0712	Pacing Problem		
								F08	Hospitalization or Prolonged Hospitalization	A070101	Failure to Capture		
										A070103	Intermittent Capture		
										A0712	Pacing Problem		
								F1203	Life Threatening Illness or Injury	A070101	Failure to Capture		
										A070103	Intermittent Capture		
										A0712	Pacing Problem		
								F23	Unexpected Medical Intervention	A070101	Failure to Capture		
										A070103	Intermittent Capture		
										A0712	Pacing Problem		
						CT233	Surgical	E0112	Dizziness	A070101	Failure to Capture		
										A070103	Intermittent Capture		
										A0712	Pacing Problem		
								E011903	Syncope/Fainting	A070101	Failure to Capture		
										A070103	Intermittent Capture		
										A0712	Pacing Problem		
								E060101	Asystole	A070101	Failure to Capture		
										A070103	Intermittent Capture		
										A0712	Pacing Problem		
								F08	Hospitalization or Prolonged Hospitalization	A070101	Failure to Capture		
										A070103	Intermittent Capture		
										A0712	Pacing Problem		
								F1203	Life Threatening Illness or Injury	A070101	Failure to Capture		
										A070103	Intermittent Capture		
										A0712	Pacing Problem		
								F23	Unexpected Medical Intervention	A070101	Failure to Capture		
										A070103	Intermittent Capture		
										A0712	Pacing Problem		
						CT2406	Implantable	E0112	Dizziness	A070101	Failure to Capture		
										A070103	Intermittent Capture		
										A0712	Pacing Problem		
								E011903	Syncope/Fainting	A070101	Failure to Capture		
										A070103	Intermittent Capture		
										A0712	Pacing Problem		
								E060101	Asystole	A070101	Failure to Capture		
										A070103	Intermittent Capture		
										A0712	Pacing Problem		
								F08	Hospitalization or Prolonged Hospitalization	A070101	Failure to Capture		
										A070103	Intermittent Capture		
										A0712	Pacing Problem		
								F1203	Life Threatening Illness or Injury	A070101	Failure to Capture		
										A070103	Intermittent Capture		
										A0712	Pacing Problem		
								F23	Unexpected Medical Intervention	A070101	Failure to Capture		
										A070103	Intermittent Capture		
										A0712	Pacing Problem		
						CT2407	Active implantable	E0112	Dizziness	A070101	Failure to Capture		
										A070103	Intermittent Capture		
										A0712	Pacing Problem		
								E011903	Syncope/Fainting	A070101	Failure to Capture		

									A070103	Intermittent Capture		
									A0712	Pacing Problem		
							E060101	Asystole	A070101	Failure to Capture		
									A070103	Intermittent Capture		
									A0712	Pacing Problem		
							F08	Hospitalization or Prolonged Hospitalization	A070101	Failure to Capture		
									A070103	Intermittent Capture		
									A0712	Pacing Problem		
							F1203	Life Threatening Illness or Injury	A070101	Failure to Capture		
									A070103	Intermittent Capture		
									A0712	Pacing Problem		
							F23	Unexpected Medical Intervention	A070101	Failure to Capture		
									A070103	Intermittent Capture		
									A0712	Pacing Problem		
					CT257	Implantable cardiac pacemakers	E0112	Dizziness	A070101	Failure to Capture		
									A070103	Intermittent Capture		
									A0712	Pacing Problem		
							E011903	Syncope/Fainting	A070101	Failure to Capture		
									A070103	Intermittent Capture		
									A0712	Pacing Problem		
							E060101	Asystole	A070101	Failure to Capture		
									A070103	Intermittent Capture		
									A0712	Pacing Problem		
							F08	Hospitalization or Prolonged Hospitalization	A070101	Failure to Capture		
									A070103	Intermittent Capture		
									A0712	Pacing Problem		
							F1203	Life Threatening Illness or Injury	A070101	Failure to Capture		
									A070103	Intermittent Capture		
									A0712	Pacing Problem		
							F23	Unexpected Medical Intervention	A070101	Failure to Capture		
									A070103	Intermittent Capture		
									A0712	Pacing Problem		
					CT2727	Active implant devices/ancillaries	E0112	Dizziness	A070101	Failure to Capture		
									A070103	Intermittent Capture		
									A0712	Pacing Problem		
							E011903	Syncope/Fainting	A070101	Failure to Capture		
									A070103	Intermittent Capture		
									A0712	Pacing Problem		
							E060101	Asystole	A070101	Failure to Capture		
									A070103	Intermittent Capture		
									A0712	Pacing Problem		
							F08	Hospitalization or Prolonged Hospitalization	A070101	Failure to Capture		
									A070103	Intermittent Capture		
									A0712	Pacing Problem		
							F1203	Life Threatening Illness or Injury	A070101	Failure to Capture		
									A070103	Intermittent Capture		
									A0712	Pacing Problem		
							F23	Unexpected Medical Intervention	A070101	Failure to Capture		
									A070103	Intermittent Capture		
									A0712	Pacing Problem		
					CT2750	Implantable cardiac electrical stimulators and associated devices	E0112	Dizziness	A070101	Failure to Capture		
									A070103	Intermittent Capture		
									A0712	Pacing Problem		
							E011903	Syncope/Fainting	A070101	Failure to Capture		
									A070103	Intermittent Capture		
									A0712	Pacing Problem		
							E060101	Asystole	A070101	Failure to Capture		
									A070103	Intermittent Capture		
									A0712	Pacing Problem		
							F08	Hospitalization or Prolonged Hospitalization	A070101	Failure to Capture		
									A070103	Intermittent Capture		
									A0712	Pacing Problem		
							F1203	Life Threatening Illness or Injury	A070101	Failure to Capture		
									A070103	Intermittent Capture		
									A0712	Pacing Problem		
							F23	Unexpected Medical Intervention	A070101	Failure to Capture		
									A070103	Intermittent Capture		
									A0712	Pacing Problem		
					CT2866	Implantable cardiac electrical stimulators	E0112	Dizziness	A070101	Failure to Capture		
									A070103	Intermittent Capture		
									A0712	Pacing Problem		
							E011903	Syncope/Fainting	A070101	Failure to Capture		
									A070103	Intermittent Capture		
									A0712	Pacing Problem		
							E060101	Asystole	A070101	Failure to Capture		
									A070103	Intermittent Capture		
									A0712	Pacing Problem		
							F08	Hospitalization or Prolonged Hospitalization	A070101	Failure to Capture		
									A070103	Intermittent Capture		
									A0712	Pacing Problem		
							F1203	Life Threatening Illness or Injury	A070101	Failure to Capture		
									A070103	Intermittent Capture		
									A0712	Pacing Problem		
							F23	Unexpected Medical Intervention	A070101	Failure to Capture		
									A070103	Intermittent Capture		
									A0712	Pacing Problem		
					CT2985	Active	E0112	Dizziness	A070101	Failure to Capture		
									A070103	Intermittent Capture		

								A0712	Pacing Problem		
						E011903	Syncope/Fainting	A070101	Failure to Capture		
								A070103	Intermittent Capture		
								A0712	Pacing Problem		
						E060101	Asystole	A070101	Failure to Capture		
								A070103	Intermittent Capture		
								A0712	Pacing Problem		
						F08	Hospitalization or Prolonged Hospitalization	A070101	Failure to Capture		
								A070103	Intermittent Capture		
								A0712	Pacing Problem		
						F1203	Life Threatening Illness or Injury	A070101	Failure to Capture		
								A070103	Intermittent Capture		
								A0712	Pacing Problem		
						F23	Unexpected Medical Intervention	A070101	Failure to Capture		
								A070103	Intermittent Capture		
								A0712	Pacing Problem		
					CT3001	Implantable metal	E0112	Dizziness	A070101	Failure to Capture		
								A070103	Intermittent Capture		
								A0712	Pacing Problem		
						E011903	Syncope/Fainting	A070101	Failure to Capture		
								A070103	Intermittent Capture		
								A0712	Pacing Problem		
						E060101	Asystole	A070101	Failure to Capture		
								A070103	Intermittent Capture		
								A0712	Pacing Problem		
						F08	Hospitalization or Prolonged Hospitalization	A070101	Failure to Capture		
								A070103	Intermittent Capture		
								A0712	Pacing Problem		
						F1203	Life Threatening Illness or Injury	A070101	Failure to Capture		
								A070103	Intermittent Capture		
								A0712	Pacing Problem		
						F23	Unexpected Medical Intervention	A070101	Failure to Capture		
								A070103	Intermittent Capture		
								A0712	Pacing Problem		
					CT3002	Implantable synthetic polymer	E0112	Dizziness	A070101	Failure to Capture		
								A070103	Intermittent Capture		
								A0712	Pacing Problem		
						E011903	Syncope/Fainting	A070101	Failure to Capture		
								A070103	Intermittent Capture		
								A0712	Pacing Problem		
						E060101	Asystole	A070101	Failure to Capture		
								A070103	Intermittent Capture		
								A0712	Pacing Problem		
						F08	Hospitalization or Prolonged Hospitalization	A070101	Failure to Capture		
								A070103	Intermittent Capture		
								A0712	Pacing Problem		
						F1203	Life Threatening Illness or Injury	A070101	Failure to Capture		
								A070103	Intermittent Capture		
								A0712	Pacing Problem		
						F23	Unexpected Medical Intervention	A070101	Failure to Capture		
								A070103	Intermittent Capture		
								A0712	Pacing Problem		
					CT3006	Implantable inorganic material	E0112	Dizziness	A070101	Failure to Capture		
								A070103	Intermittent Capture		
								A0712	Pacing Problem		
						E011903	Syncope/Fainting	A070101	Failure to Capture		
								A070103	Intermittent Capture		
								A0712	Pacing Problem		
						E060101	Asystole	A070101	Failure to Capture		
								A070103	Intermittent Capture		
								A0712	Pacing Problem		
						F08	Hospitalization or Prolonged Hospitalization	A070101	Failure to Capture		
								A070103	Intermittent Capture		
								A0712	Pacing Problem		
						F1203	Life Threatening Illness or Injury	A070101	Failure to Capture		
								A070103	Intermittent Capture		
								A0712	Pacing Problem		
						F23	Unexpected Medical Intervention	A070101	Failure to Capture		
								A070103	Intermittent Capture		
								A0712	Pacing Problem		
					CT314	Electrically powered	E0112	Dizziness	A070101	Failure to Capture		
								A070103	Intermittent Capture		
								A0712	Pacing Problem		
						E011903	Syncope/Fainting	A070101	Failure to Capture		
								A070103	Intermittent Capture		
								A0712	Pacing Problem		
						E060101	Asystole	A070101	Failure to Capture		
								A070103	Intermittent Capture		
								A0712	Pacing Problem		
						F08	Hospitalization or Prolonged Hospitalization	A070101	Failure to Capture		
								A070103	Intermittent Capture		
								A0712	Pacing Problem		
						F1203	Life Threatening Illness or Injury	A070101	Failure to Capture		
								A070103	Intermittent Capture		
								A0712	Pacing Problem		
						F23	Unexpected Medical Intervention	A070101	Failure to Capture		
								A070103	Intermittent Capture		
								A0712	Pacing Problem		

					CT321	Long-term surgical invasive	E0112	Dizziness	A070101	Failure to Capture		
									A070103	Intermittent Capture		
									A0712	Pacing Problem		
							E011903	Syncope/Fainting	A070101	Failure to Capture		
									A070103	Intermittent Capture		
									A0712	Pacing Problem		
							E060101	Asystole	A070101	Failure to Capture		
									A070103	Intermittent Capture		
									A0712	Pacing Problem		
							F08	Hospitalization or Prolonged Hospitalization	A070101	Failure to Capture		
									A070103	Intermittent Capture		
									A0712	Pacing Problem		
							F1203	Life Threatening Illness or Injury	A070101	Failure to Capture		
									A070103	Intermittent Capture		
									A0712	Pacing Problem		
							F23	Unexpected Medical Intervention	A070101	Failure to Capture		
									A070103	Intermittent Capture		
									A0712	Pacing Problem		
					CT449	Generators and associated devices	E0112	Dizziness	A070101	Failure to Capture		
									A070103	Intermittent Capture		
									A0712	Pacing Problem		
							E011903	Syncope/Fainting	A070101	Failure to Capture		
									A070103	Intermittent Capture		
									A0712	Pacing Problem		
							E060101	Asystole	A070101	Failure to Capture		
									A070103	Intermittent Capture		
									A0712	Pacing Problem		
							F08	Hospitalization or Prolonged Hospitalization	A070101	Failure to Capture		
									A070103	Intermittent Capture		
									A0712	Pacing Problem		
							F1203	Life Threatening Illness or Injury	A070101	Failure to Capture		
									A070103	Intermittent Capture		
									A0712	Pacing Problem		
							F23	Unexpected Medical Intervention	A070101	Failure to Capture		
									A070103	Intermittent Capture		
									A0712	Pacing Problem		
					CT981	Single use	E0112	Dizziness	A070101	Failure to Capture		
									A070103	Intermittent Capture		
									A0712	Pacing Problem		
							E011903	Syncope/Fainting	A070101	Failure to Capture		
									A070103	Intermittent Capture		
									A0712	Pacing Problem		
							E060101	Asystole	A070101	Failure to Capture		
									A070103	Intermittent Capture		
									A0712	Pacing Problem		
							F08	Hospitalization or Prolonged Hospitalization	A070101	Failure to Capture		
									A070103	Intermittent Capture		
									A0712	Pacing Problem		
							F1203	Life Threatening Illness or Injury	A070101	Failure to Capture		
									A070103	Intermittent Capture		
									A0712	Pacing Problem		
							F23	Unexpected Medical Intervention	A070101	Failure to Capture		
									A070103	Intermittent Capture		
									A0712	Pacing Problem		
					CT983	Surgical invasive	E0112	Dizziness	A070101	Failure to Capture		
									A070103	Intermittent Capture		
									A0712	Pacing Problem		
							E011903	Syncope/Fainting	A070101	Failure to Capture		
									A070103	Intermittent Capture		
									A0712	Pacing Problem		
							E060101	Asystole	A070101	Failure to Capture		
									A070103	Intermittent Capture		
									A0712	Pacing Problem		
							F08	Hospitalization or Prolonged Hospitalization	A070101	Failure to Capture		
									A070103	Intermittent Capture		
									A0712	Pacing Problem		
							F1203	Life Threatening Illness or Injury	A070101	Failure to Capture		
									A070103	Intermittent Capture		
									A0712	Pacing Problem		
							F23	Unexpected Medical Intervention	A070101	Failure to Capture		
									A070103	Intermittent Capture		
									A0712	Pacing Problem		
8009	2022	All other reportable incidents	17882	Physiologic-monitoring defibrillation system	CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070803	Failure to Power Up		
							F27	Problem Identified During Non-Clinical Procedure	A070803	Failure to Power Up		
8010	2022	All other reportable incidents	35852	Single-chamber implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	12/10/2022	12/10/2022
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	12/10/2022	12/10/2022
							F12	Serious Injury/ Illness/ Impairment	A070504	Premature Discharge of Battery	12/10/2022	12/10/2022
							F1903	Device Explantation	A070504	Premature Discharge of Battery	12/10/2022	12/10/2022
8011	2022	All other reportable incidents	17882	Physiologic-monitoring defibrillation system	CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070803	Failure to Power Up		
							F27	Problem Identified During Non-Clinical Procedure	A070803	Failure to Power Up		
8012	2022	All other reportable incidents	35223	Endocardial/interventricular septal pacing lead	CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070102	High Capture Threshold		
									A070909	Oversensing		
									A0722	Impedance Problem		
							F26	No Health Consequences or Impact	A070102	High Capture Threshold		
									A070909	Oversensing		
									A0722	Impedance Problem		
8013	2022	All other reportable incidents	17882	Physiologic-monitoring defibrillation system	CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070803	Failure to Power Up		
							F27	Problem Identified During Non-Clinical Procedure	A070803	Failure to Power Up		

8014	2022	All other reportable incidents	47270	Cardiac resynchronization therapy implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A0713	Defibrillation/Stimulation Problem		
									A090811	Unable to Obtain Readings		
							F26	No Health Consequences or Impact	A0713	Defibrillation/Stimulation Problem		
									A090811	Unable to Obtain Readings		
8015	2022	All other reportable incidents	17882	Physiologic-monitoring defibrillation system	CT1684	Cardiac pacemakers and associated devices	E0602	Cardiac Arrest	A070908	Failure to Sense		
							E2403	No Clinical Signs, Symptoms or Conditions	A070908	Failure to Sense		
							F26	No Health Consequences or Impact	A070908	Failure to Sense		
8016	2022	All other reportable incidents	17882	Physiologic-monitoring defibrillation system	CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070803	Failure to Power Up		
							F27	Problem Identified During Non-Clinical Procedure	A070803	Failure to Power Up		
8017	2022	All other reportable incidents	17882	Physiologic-monitoring defibrillation system	CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070803	Failure to Power Up		
							F27	Problem Identified During Non-Clinical Procedure	A070803	Failure to Power Up		
8018	2022	All other reportable incidents	17882	Physiologic-monitoring defibrillation system	CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070803	Failure to Power Up		
							F27	Problem Identified During Non-Clinical Procedure	A070803	Failure to Power Up		
8019	2022	All other reportable incidents	36241	Pacemaker/defibrillator lead	CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070101	Failure to Capture		
									A070102	High Capture Threshold		
									A072201	High Impedance		
							F08	Hospitalization or Prolonged Hospitalization	A070101	Failure to Capture		
									A070102	High Capture Threshold		
									A072201	High Impedance		
							F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A070101	Failure to Capture		
									A070102	High Capture Threshold		
									A072201	High Impedance		
8020	2022	All other reportable incidents	17882	Physiologic-monitoring defibrillation system	CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A0801	Failure to Calibrate		
							F27	Problem Identified During Non-Clinical Procedure	A0801	Failure to Calibrate		
8021	2022	All other reportable incidents	17882	Physiologic-monitoring defibrillation system	CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070803	Failure to Power Up		
							F27	Problem Identified During Non-Clinical Procedure	A070803	Failure to Power Up		
8022	2022	All other reportable incidents	47270	Cardiac resynchronization therapy implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	19/10/2022	19/10/2022
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	19/10/2022	19/10/2022
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	19/10/2022	19/10/2022
8023	2022	All other reportable incidents	37265	Dual-chamber implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A070909	Oversensing		
									A072201	High Impedance		
							F26	No Health Consequences or Impact	A070909	Oversensing		
									A072201	High Impedance		
8024	2022	All other reportable incidents	36241	Pacemaker/defibrillator lead	CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070909	Oversensing		
									A072201	High Impedance		
							F26	No Health Consequences or Impact	A070909	Oversensing		
									A072201	High Impedance		
8025	2022	All other reportable incidents	47270	Cardiac resynchronization therapy implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
8026	2022	All other reportable incidents	35852	Single-chamber implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A070909	Oversensing		
									A1301	Failure to Read Input Signal		
							F26	No Health Consequences or Impact	A070909	Oversensing		
									A1301	Failure to Read Input Signal		
8027	2022	All other reportable incidents	35852	Single-chamber implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A070909	Oversensing	25/01/2022	25/01/2022
									A090801	Signal Artifact/Noise	25/01/2022	25/01/2022
							F1908	Prolonged Surgery	A070909	Oversensing	25/01/2022	25/01/2022
									A090801	Signal Artifact/Noise	25/01/2022	25/01/2022
8028	2022	All other reportable incidents	17882	Physiologic-monitoring defibrillation system	CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A071301	Failure to Deliver Shock/Stimulation		
							F27	Problem Identified During Non-Clinical Procedure	A071301	Failure to Deliver Shock/Stimulation		
8029	2022	All other reportable incidents	36241	Pacemaker/defibrillator lead	CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A051201	Device Dislodged or Dislocated	17/02/2022	17/02/2022
							F08	Hospitalization or Prolonged Hospitalization	A051201	Device Dislodged or Dislocated	17/02/2022	17/02/2022
							F1905	Device Revision or Replacement	A051201	Device Dislodged or Dislocated	17/02/2022	17/02/2022
8030	2022	All other reportable incidents	35224	External pacemaker, epicardial pacing	CT1684	Cardiac pacemakers and associated devices	E060110	Ventricular Fibrillation	A0701	Capturing Problem		
							E2403	No Clinical Signs, Symptoms or Conditions	A0701	Capturing Problem		
							F12	Serious Injury/ Illness/ Impairment	A0701	Capturing Problem		
					CT1685	Cardiac pacemakers	E060110	Ventricular Fibrillation	A0701	Capturing Problem		
							E2403	No Clinical Signs, Symptoms or Conditions	A0701	Capturing Problem		
							F12	Serious Injury/ Illness/ Impairment	A0701	Capturing Problem		
8031	2022	All other reportable incidents	35852	Single-chamber implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem	16/03/2022	16/03/2022
									A070504	Premature Discharge of Battery	16/03/2022	16/03/2022
							F08	Hospitalization or Prolonged Hospitalization	A0705	Battery Problem	16/03/2022	16/03/2022
									A070504	Premature Discharge of Battery	16/03/2022	16/03/2022
							F1905	Device Revision or Replacement	A0705	Battery Problem	16/03/2022	16/03/2022
									A070504	Premature Discharge of Battery	16/03/2022	16/03/2022
8032	2022	All other reportable incidents	35852	Single-chamber implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem	31/03/2022	31/03/2022
									A070504	Premature Discharge of Battery	31/03/2022	31/03/2022
							F08	Hospitalization or Prolonged Hospitalization	A0705	Battery Problem	31/03/2022	31/03/2022
									A070504	Premature Discharge of Battery	31/03/2022	31/03/2022
							F1905	Device Revision or Replacement	A0705	Battery Problem	31/03/2022	31/03/2022
									A070504	Premature Discharge of Battery	31/03/2022	31/03/2022
8033	2022	All other reportable incidents	17882	Physiologic-monitoring defibrillation system	CT1684	Cardiac pacemakers and associated devices	E0602	Cardiac Arrest	A071301	Failure to Deliver Shock/Stimulation		
							E2403	No Clinical Signs, Symptoms or Conditions	A071301	Failure to Deliver Shock/Stimulation		
							F26	No Health Consequences or Impact	A071301	Failure to Deliver Shock/Stimulation		
8034	2022	All other reportable incidents	47267	Single-chamber implantable pacemaker, rate-responsive	CT141	Electrodes/Leads and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
					CT145	Cardiovascular devices	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
					CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
					CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A0705	Battery Problem		

									A070504	Premature Discharge of Battery		
					CT233	Surgical	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
					CT2406	Implantable	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
					CT2407	Active implantable	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
					CT2727	Active implant devices/ancillaries	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
					CT2750	Implantable cardiac electrical stimulators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
					CT2985	Active	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
					CT3001	Implantable metal	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
					CT3002	Implantable synthetic polymer	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
					CT3006	Implantable inorganic material	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
					CT314	Electrically powered	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
					CT321	Long-term surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
					CT449	Generators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
					CT981	Single use	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
					CT983	Surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
8035	2022	All other reportable incidents	17882	Physiologic-monitoring defibrillation system	CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A071302	Inappropriate/Inadequate Shock/Stimulation		
							F26	No Health Consequences or Impact	A071302	Inappropriate/Inadequate Shock/Stimulation		
8036	2022	All other reportable incidents	17882	Physiologic-monitoring defibrillation system	CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070803	Failure to Power Up		
							F27	Problem Identified During Non-Clinical Procedure	A070803	Failure to Power Up		
8037	2022	All other reportable incidents	17882	Physiologic-monitoring defibrillation system	CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A0719	Unexpected Shutdown		
							F26	No Health Consequences or Impact	A0719	Unexpected Shutdown		
8038	2022	Death	57968	Cardiovascular device introducer, steerable	CT1684	Cardiac pacemakers and associated devices	E2401	Insufficient Information	A040506	Peeled/Delaminated		
									A0414	Material Split, Cut or Torn		
							F02	Death	A040506	Peeled/Delaminated		
									A0414	Material Split, Cut or Torn		
							F1901	Additional Surgery	A040506	Peeled/Delaminated		
									A0414	Material Split, Cut or Torn		
8039	2022	All other reportable incidents	47205	Cardiac pulse generator programmer	CT1684	Cardiac pacemakers and associated devices	E060101	Asystole	A0712	Pacing Problem		
									A11	Computer Software Problem		
									A1104	Computer Operating System Problem		
							E2403	No Clinical Signs, Symptoms or Conditions	A0712	Pacing Problem		
									A11	Computer Software Problem		
									A1104	Computer Operating System Problem		
							F23	Unexpected Medical Intervention	A0712	Pacing Problem		
									A11	Computer Software Problem		
									A1104	Computer Operating System Problem		
8040	2022	All other reportable incidents	47265	Dual-chamber implantable pacemaker, rate-responsive	CT141	Electrodes/Leads and associated devices	E0204	Lethargy	A070504	Premature Discharge of Battery		
									A090804	False Positive Result		

		E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
				A090804	False Positive Result		
		F11	Non-Serious Injury/ Illness/ Impairment	A070504	Premature Discharge of Battery		
				A090804	False Positive Result		
CT145	Cardiovascular devices	E0204	Lethargy	A070504	Premature Discharge of Battery		
				A090804	False Positive Result		
		E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
				A090804	False Positive Result		
		F11	Non-Serious Injury/ Illness/ Impairment	A070504	Premature Discharge of Battery		
				A090804	False Positive Result		
CT1684	Cardiac pacemakers and associated devices	E0204	Lethargy	A070504	Premature Discharge of Battery		
				A090804	False Positive Result		
		E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
				A090804	False Positive Result		
		F11	Non-Serious Injury/ Illness/ Impairment	A070504	Premature Discharge of Battery		
				A090804	False Positive Result		
CT1685	Cardiac pacemakers	E0204	Lethargy	A070504	Premature Discharge of Battery		
				A090804	False Positive Result		
		E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
				A090804	False Positive Result		
		F11	Non-Serious Injury/ Illness/ Impairment	A070504	Premature Discharge of Battery		
				A090804	False Positive Result		
CT233	Surgical	E0204	Lethargy	A070504	Premature Discharge of Battery		
				A090804	False Positive Result		
		E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
				A090804	False Positive Result		
		F11	Non-Serious Injury/ Illness/ Impairment	A070504	Premature Discharge of Battery		
				A090804	False Positive Result		
CT2406	Implantable	E0204	Lethargy	A070504	Premature Discharge of Battery		
				A090804	False Positive Result		
		E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
				A090804	False Positive Result		
		F11	Non-Serious Injury/ Illness/ Impairment	A070504	Premature Discharge of Battery		
				A090804	False Positive Result		
CT2407	Active implantable	E0204	Lethargy	A070504	Premature Discharge of Battery		
				A090804	False Positive Result		
		E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
				A090804	False Positive Result		
		F11	Non-Serious Injury/ Illness/ Impairment	A070504	Premature Discharge of Battery		
				A090804	False Positive Result		
CT257	Implantable cardiac pacemakers	E0204	Lethargy	A070504	Premature Discharge of Battery		
				A090804	False Positive Result		
		E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
				A090804	False Positive Result		
		F11	Non-Serious Injury/ Illness/ Impairment	A070504	Premature Discharge of Battery		
				A090804	False Positive Result		
CT2727	Active implant devices/ancillaries	E0204	Lethargy	A070504	Premature Discharge of Battery		
				A090804	False Positive Result		
		E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
				A090804	False Positive Result		
		F11	Non-Serious Injury/ Illness/ Impairment	A070504	Premature Discharge of Battery		
				A090804	False Positive Result		
CT2750	Implantable cardiac electrical stimulators and associated devices	E0204	Lethargy	A070504	Premature Discharge of Battery		
				A090804	False Positive Result		
		E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
				A090804	False Positive Result		
		F11	Non-Serious Injury/ Illness/ Impairment	A070504	Premature Discharge of Battery		
				A090804	False Positive Result		
CT2866	Implantable cardiac electrical stimulators	E0204	Lethargy	A070504	Premature Discharge of Battery		
				A090804	False Positive Result		
		E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
				A090804	False Positive Result		
		F11	Non-Serious Injury/ Illness/ Impairment	A070504	Premature Discharge of Battery		
				A090804	False Positive Result		
CT2985	Active	E0204	Lethargy	A070504	Premature Discharge of Battery		
				A090804	False Positive Result		
		E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
				A090804	False Positive Result		
		F11	Non-Serious Injury/ Illness/ Impairment	A070504	Premature Discharge of Battery		
				A090804	False Positive Result		
CT3001	Implantable metal	E0204	Lethargy	A070504	Premature Discharge of Battery		
				A090804	False Positive Result		
		E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
				A090804	False Positive Result		
		F11	Non-Serious Injury/ Illness/ Impairment	A070504	Premature Discharge of Battery		
				A090804	False Positive Result		
CT3002	Implantable synthetic polymer	E0204	Lethargy	A070504	Premature Discharge of Battery		
				A090804	False Positive Result		
		E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
				A090804	False Positive Result		
		F11	Non-Serious Injury/ Illness/ Impairment	A070504	Premature Discharge of Battery		
				A090804	False Positive Result		
CT3006	Implantable inorganic material	E0204	Lethargy	A070504	Premature Discharge of Battery		
				A090804	False Positive Result		
		E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
				A090804	False Positive Result		
		F11	Non-Serious Injury/ Illness/ Impairment	A070504	Premature Discharge of Battery		
				A090804	False Positive Result		

					CT314	Electrically powered	E0204	Lethargy	A070504	Premature Discharge of Battery		
									A090804	False Positive Result		
							E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
									A090804	False Positive Result		
							F11	Non-Serious Injury/ Illness/ Impairment	A070504	Premature Discharge of Battery		
									A090804	False Positive Result		
					CT321	Long-term surgical invasive	E0204	Lethargy	A070504	Premature Discharge of Battery		
									A090804	False Positive Result		
							E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
									A090804	False Positive Result		
							F11	Non-Serious Injury/ Illness/ Impairment	A070504	Premature Discharge of Battery		
									A090804	False Positive Result		
					CT449	Generators and associated devices	E0204	Lethargy	A070504	Premature Discharge of Battery		
									A090804	False Positive Result		
							E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
									A090804	False Positive Result		
							F11	Non-Serious Injury/ Illness/ Impairment	A070504	Premature Discharge of Battery		
									A090804	False Positive Result		
					CT981	Single use	E0204	Lethargy	A070504	Premature Discharge of Battery		
									A090804	False Positive Result		
							E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
									A090804	False Positive Result		
							F11	Non-Serious Injury/ Illness/ Impairment	A070504	Premature Discharge of Battery		
									A090804	False Positive Result		
					CT983	Surgical invasive	E0204	Lethargy	A070504	Premature Discharge of Battery		
									A090804	False Positive Result		
							E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
									A090804	False Positive Result		
							F11	Non-Serious Injury/ Illness/ Impairment	A070504	Premature Discharge of Battery		
									A090804	False Positive Result		
8041	2022	All other reportable incidents	47270	Cardiac resynchronization therapy implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
8042	2022	All other reportable incidents	47270	Cardiac resynchronization therapy implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	11/02/2022	11/02/2022
							F1901	Additional Surgery	A070504	Premature Discharge of Battery	11/02/2022	11/02/2022
8043	2022	All other reportable incidents	35039	Epicardial pacing lead	CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070909	Oversensing	19/08/2022	19/08/2022
							F08	Hospitalization or Prolonged Hospitalization	A070909	Oversensing	19/08/2022	19/08/2022
							F12	Serious Injury/ Illness/ Impairment	A070909	Oversensing	19/08/2022	19/08/2022
							F1905	Device Revision or Replacement	A070909	Oversensing	19/08/2022	19/08/2022
							F23	Unexpected Medical Intervention	A070909	Oversensing	19/08/2022	19/08/2022
8044	2022	All other reportable incidents	37265	Dual-chamber implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery		
							F12	Serious Injury/ Illness/ Impairment	A070504	Premature Discharge of Battery		
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery		
8045	2022	All other reportable incidents	17882	Physiologic-monitoring defibrillation system	CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A071302	Inappropriate/Inadequate Shock/Stimulation		
							F26	No Health Consequences or Impact	A071302	Inappropriate/Inadequate Shock/Stimulation		
8046	2022	All other reportable incidents	47267	Single-chamber implantable pacemaker, rate-responsive	CT141	Electrodes/Leads and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A0203	Defective Device	20/07/2022	20/07/2022
									A0705	Battery Problem	20/07/2022	20/07/2022
									A070504	Premature Discharge of Battery	20/07/2022	20/07/2022
							F08	Hospitalization or Prolonged Hospitalization	A0203	Defective Device	20/07/2022	20/07/2022
									A0705	Battery Problem	20/07/2022	20/07/2022
									A070504	Premature Discharge of Battery	20/07/2022	20/07/2022
							F1905	Device Revision or Replacement	A0203	Defective Device	20/07/2022	20/07/2022
									A0705	Battery Problem	20/07/2022	20/07/2022
									A070504	Premature Discharge of Battery	20/07/2022	20/07/2022
					CT145	Cardiovascular devices	E2403	No Clinical Signs, Symptoms or Conditions	A0203	Defective Device	20/07/2022	20/07/2022
									A0705	Battery Problem	20/07/2022	20/07/2022
									A070504	Premature Discharge of Battery	20/07/2022	20/07/2022
							F08	Hospitalization or Prolonged Hospitalization	A0203	Defective Device	20/07/2022	20/07/2022
									A0705	Battery Problem	20/07/2022	20/07/2022
									A070504	Premature Discharge of Battery	20/07/2022	20/07/2022
							F1905	Device Revision or Replacement	A0203	Defective Device	20/07/2022	20/07/2022
									A0705	Battery Problem	20/07/2022	20/07/2022
									A070504	Premature Discharge of Battery	20/07/2022	20/07/2022
					CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A0203	Defective Device	20/07/2022	20/07/2022
									A0705	Battery Problem	20/07/2022	20/07/2022
									A070504	Premature Discharge of Battery	20/07/2022	20/07/2022
							F08	Hospitalization or Prolonged Hospitalization	A0203	Defective Device	20/07/2022	20/07/2022
									A0705	Battery Problem	20/07/2022	20/07/2022
									A070504	Premature Discharge of Battery	20/07/2022	20/07/2022
							F1905	Device Revision or Replacement	A0203	Defective Device	20/07/2022	20/07/2022
									A0705	Battery Problem	20/07/2022	20/07/2022
									A070504	Premature Discharge of Battery	20/07/2022	20/07/2022
					CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A0203	Defective Device	20/07/2022	20/07/2022
									A0705	Battery Problem	20/07/2022	20/07/2022
									A070504	Premature Discharge of Battery	20/07/2022	20/07/2022
							F08	Hospitalization or Prolonged Hospitalization	A0203	Defective Device	20/07/2022	20/07/2022
									A0705	Battery Problem	20/07/2022	20/07/2022
									A070504	Premature Discharge of Battery	20/07/2022	20/07/2022
							F1905	Device Revision or Replacement	A0203	Defective Device	20/07/2022	20/07/2022
									A0705	Battery Problem	20/07/2022	20/07/2022
									A070504	Premature Discharge of Battery	20/07/2022	20/07/2022
					CT233	Surgical	E2403	No Clinical Signs, Symptoms or Conditions	A0203	Defective Device	20/07/2022	20/07/2022
									A0705	Battery Problem	20/07/2022	20/07/2022
									A070504	Premature Discharge of Battery	20/07/2022	20/07/2022
							F08	Hospitalization or Prolonged Hospitalization	A0203	Defective Device	20/07/2022	20/07/2022
									A0705	Battery Problem	20/07/2022	20/07/2022
									A070504	Premature Discharge of Battery	20/07/2022	20/07/2022
							F1905	Device Revision or Replacement	A0203	Defective Device	20/07/2022	20/07/2022

								A0705	Battery Problem	20/07/2022	20/07/2022
								A070504	Premature Discharge of Battery	20/07/2022	20/07/2022
					CT2406	Implantable	E2403	No Clinical Signs, Symptoms or Conditions	A0203	Defective Device	20/07/2022	20/07/2022
									A0705	Battery Problem	20/07/2022	20/07/2022
									A070504	Premature Discharge of Battery	20/07/2022	20/07/2022
							F08	Hospitalization or Prolonged Hospitalization	A0203	Defective Device	20/07/2022	20/07/2022
									A0705	Battery Problem	20/07/2022	20/07/2022
									A070504	Premature Discharge of Battery	20/07/2022	20/07/2022
							F1905	Device Revision or Replacement	A0203	Defective Device	20/07/2022	20/07/2022
									A0705	Battery Problem	20/07/2022	20/07/2022
									A070504	Premature Discharge of Battery	20/07/2022	20/07/2022
					CT2407	Active implantable	E2403	No Clinical Signs, Symptoms or Conditions	A0203	Defective Device	20/07/2022	20/07/2022
									A0705	Battery Problem	20/07/2022	20/07/2022
									A070504	Premature Discharge of Battery	20/07/2022	20/07/2022
							F08	Hospitalization or Prolonged Hospitalization	A0203	Defective Device	20/07/2022	20/07/2022
									A0705	Battery Problem	20/07/2022	20/07/2022
									A070504	Premature Discharge of Battery	20/07/2022	20/07/2022
							F1905	Device Revision or Replacement	A0203	Defective Device	20/07/2022	20/07/2022
									A0705	Battery Problem	20/07/2022	20/07/2022
									A070504	Premature Discharge of Battery	20/07/2022	20/07/2022
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A0203	Defective Device	20/07/2022	20/07/2022
									A0705	Battery Problem	20/07/2022	20/07/2022
									A070504	Premature Discharge of Battery	20/07/2022	20/07/2022
							F08	Hospitalization or Prolonged Hospitalization	A0203	Defective Device	20/07/2022	20/07/2022
									A0705	Battery Problem	20/07/2022	20/07/2022
									A070504	Premature Discharge of Battery	20/07/2022	20/07/2022
							F1905	Device Revision or Replacement	A0203	Defective Device	20/07/2022	20/07/2022
									A0705	Battery Problem	20/07/2022	20/07/2022
									A070504	Premature Discharge of Battery	20/07/2022	20/07/2022
					CT2727	Active implant devices/ancillaries	E2403	No Clinical Signs, Symptoms or Conditions	A0203	Defective Device	20/07/2022	20/07/2022
									A0705	Battery Problem	20/07/2022	20/07/2022
									A070504	Premature Discharge of Battery	20/07/2022	20/07/2022
							F08	Hospitalization or Prolonged Hospitalization	A0203	Defective Device	20/07/2022	20/07/2022
									A0705	Battery Problem	20/07/2022	20/07/2022
									A070504	Premature Discharge of Battery	20/07/2022	20/07/2022
							F1905	Device Revision or Replacement	A0203	Defective Device	20/07/2022	20/07/2022
									A0705	Battery Problem	20/07/2022	20/07/2022
									A070504	Premature Discharge of Battery	20/07/2022	20/07/2022
					CT2750	Implantable cardiac electrical stimulators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A0203	Defective Device	20/07/2022	20/07/2022
									A0705	Battery Problem	20/07/2022	20/07/2022
									A070504	Premature Discharge of Battery	20/07/2022	20/07/2022
							F08	Hospitalization or Prolonged Hospitalization	A0203	Defective Device	20/07/2022	20/07/2022
									A0705	Battery Problem	20/07/2022	20/07/2022
									A070504	Premature Discharge of Battery	20/07/2022	20/07/2022
							F1905	Device Revision or Replacement	A0203	Defective Device	20/07/2022	20/07/2022
									A0705	Battery Problem	20/07/2022	20/07/2022
									A070504	Premature Discharge of Battery	20/07/2022	20/07/2022
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A0203	Defective Device	20/07/2022	20/07/2022
									A0705	Battery Problem	20/07/2022	20/07/2022
									A070504	Premature Discharge of Battery	20/07/2022	20/07/2022
							F08	Hospitalization or Prolonged Hospitalization	A0203	Defective Device	20/07/2022	20/07/2022
									A0705	Battery Problem	20/07/2022	20/07/2022
									A070504	Premature Discharge of Battery	20/07/2022	20/07/2022
							F1905	Device Revision or Replacement	A0203	Defective Device	20/07/2022	20/07/2022
									A0705	Battery Problem	20/07/2022	20/07/2022
									A070504	Premature Discharge of Battery	20/07/2022	20/07/2022
					CT2985	Active	E2403	No Clinical Signs, Symptoms or Conditions	A0203	Defective Device	20/07/2022	20/07/2022
									A0705	Battery Problem	20/07/2022	20/07/2022
									A070504	Premature Discharge of Battery	20/07/2022	20/07/2022
							F08	Hospitalization or Prolonged Hospitalization	A0203	Defective Device	20/07/2022	20/07/2022
									A0705	Battery Problem	20/07/2022	20/07/2022
									A070504	Premature Discharge of Battery	20/07/2022	20/07/2022
							F1905	Device Revision or Replacement	A0203	Defective Device	20/07/2022	20/07/2022
									A0705	Battery Problem	20/07/2022	20/07/2022
									A070504	Premature Discharge of Battery	20/07/2022	20/07/2022
					CT3001	Implantable metal	E2403	No Clinical Signs, Symptoms or Conditions	A0203	Defective Device	20/07/2022	20/07/2022
									A0705	Battery Problem	20/07/2022	20/07/2022
									A070504	Premature Discharge of Battery	20/07/2022	20/07/2022
							F08	Hospitalization or Prolonged Hospitalization	A0203	Defective Device	20/07/2022	20/07/2022
									A0705	Battery Problem	20/07/2022	20/07/2022
									A070504	Premature Discharge of Battery	20/07/2022	20/07/2022
							F1905	Device Revision or Replacement	A0203	Defective Device	20/07/2022	20/07/2022
									A0705	Battery Problem	20/07/2022	20/07/2022
									A070504	Premature Discharge of Battery	20/07/2022	20/07/2022
					CT3002	Implantable synthetic polymer	E2403	No Clinical Signs, Symptoms or Conditions	A0203	Defective Device	20/07/2022	20/07/2022
									A0705	Battery Problem	20/07/2022	20/07/2022
									A070504	Premature Discharge of Battery	20/07/2022	20/07/2022
							F08	Hospitalization or Prolonged Hospitalization	A0203	Defective Device	20/07/2022	20/07/2022
									A0705	Battery Problem	20/07/2022	20/07/2022
									A070504	Premature Discharge of Battery	20/07/2022	20/07/2022
							F1905	Device Revision or Replacement	A0203	Defective Device	20/07/2022	20/07/2022
									A0705	Battery Problem	20/07/2022	20/07/2022
									A070504	Premature Discharge of Battery	20/07/2022	20/07/2022
					CT3006	Implantable inorganic material	E2403	No Clinical Signs, Symptoms or Conditions	A0203	Defective Device	20/07/2022	20/07/2022
									A0705	Battery Problem	20/07/2022	20/07/2022
									A070504	Premature Discharge of Battery	20/07/2022	20/07/2022
							F08	Hospitalization or Prolonged Hospitalization	A0203	Defective Device	20/07/2022	20/07/2022
									A0705	Battery Problem	20/07/2022	20/07/2022

									A070504	Premature Discharge of Battery	20/07/2022	20/07/2022
							F1905	Device Revision or Replacement	A0203	Defective Device	20/07/2022	20/07/2022
									A0705	Battery Problem	20/07/2022	20/07/2022
									A070504	Premature Discharge of Battery	20/07/2022	20/07/2022
					CT314	Electrically powered	E2403	No Clinical Signs, Symptoms or Conditions	A0203	Defective Device	20/07/2022	20/07/2022
									A0705	Battery Problem	20/07/2022	20/07/2022
									A070504	Premature Discharge of Battery	20/07/2022	20/07/2022
							F08	Hospitalization or Prolonged Hospitalization	A0203	Defective Device	20/07/2022	20/07/2022
									A0705	Battery Problem	20/07/2022	20/07/2022
									A070504	Premature Discharge of Battery	20/07/2022	20/07/2022
							F1905	Device Revision or Replacement	A0203	Defective Device	20/07/2022	20/07/2022
									A0705	Battery Problem	20/07/2022	20/07/2022
									A070504	Premature Discharge of Battery	20/07/2022	20/07/2022
					CT321	Long-term surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A0203	Defective Device	20/07/2022	20/07/2022
									A0705	Battery Problem	20/07/2022	20/07/2022
									A070504	Premature Discharge of Battery	20/07/2022	20/07/2022
							F08	Hospitalization or Prolonged Hospitalization	A0203	Defective Device	20/07/2022	20/07/2022
									A0705	Battery Problem	20/07/2022	20/07/2022
									A070504	Premature Discharge of Battery	20/07/2022	20/07/2022
							F1905	Device Revision or Replacement	A0203	Defective Device	20/07/2022	20/07/2022
									A0705	Battery Problem	20/07/2022	20/07/2022
									A070504	Premature Discharge of Battery	20/07/2022	20/07/2022
					CT449	Generators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A0203	Defective Device	20/07/2022	20/07/2022
									A0705	Battery Problem	20/07/2022	20/07/2022
									A070504	Premature Discharge of Battery	20/07/2022	20/07/2022
							F08	Hospitalization or Prolonged Hospitalization	A0203	Defective Device	20/07/2022	20/07/2022
									A0705	Battery Problem	20/07/2022	20/07/2022
									A070504	Premature Discharge of Battery	20/07/2022	20/07/2022
							F1905	Device Revision or Replacement	A0203	Defective Device	20/07/2022	20/07/2022
									A0705	Battery Problem	20/07/2022	20/07/2022
									A070504	Premature Discharge of Battery	20/07/2022	20/07/2022
					CT981	Single use	E2403	No Clinical Signs, Symptoms or Conditions	A0203	Defective Device	20/07/2022	20/07/2022
									A0705	Battery Problem	20/07/2022	20/07/2022
									A070504	Premature Discharge of Battery	20/07/2022	20/07/2022
							F08	Hospitalization or Prolonged Hospitalization	A0203	Defective Device	20/07/2022	20/07/2022
									A0705	Battery Problem	20/07/2022	20/07/2022
									A070504	Premature Discharge of Battery	20/07/2022	20/07/2022
							F1905	Device Revision or Replacement	A0203	Defective Device	20/07/2022	20/07/2022
									A0705	Battery Problem	20/07/2022	20/07/2022
									A070504	Premature Discharge of Battery	20/07/2022	20/07/2022
					CT983	Surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A0203	Defective Device	20/07/2022	20/07/2022
									A0705	Battery Problem	20/07/2022	20/07/2022
									A070504	Premature Discharge of Battery	20/07/2022	20/07/2022
							F08	Hospitalization or Prolonged Hospitalization	A0203	Defective Device	20/07/2022	20/07/2022
									A0705	Battery Problem	20/07/2022	20/07/2022
									A070504	Premature Discharge of Battery	20/07/2022	20/07/2022
							F1905	Device Revision or Replacement	A0203	Defective Device	20/07/2022	20/07/2022
									A0705	Battery Problem	20/07/2022	20/07/2022
									A070504	Premature Discharge of Battery	20/07/2022	20/07/2022
8047	2022	All other reportable incidents	47270	Cardiac resynchronization therapy implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A072201	High Impedance		
							F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A072201	High Impedance		
8048	2022	All other reportable incidents	36241	Pacemaker/defibrillator lead	CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A072201	High Impedance		
							F26	No Health Consequences or Impact	A072201	High Impedance		
8049	2022	All other reportable incidents	35223	Endocardial/interventricular septal pacing lead	CT1684	Cardiac pacemakers and associated devices	E060106	Heart Block	A051201	Device Dislodged or Dislocated	05/10/2022	05/10/2022
									A24	No Apparent Device Problem	05/10/2022	05/10/2022
							E0602	Cardiac Arrest	A051201	Device Dislodged or Dislocated	05/10/2022	05/10/2022
									A24	No Apparent Device Problem	05/10/2022	05/10/2022
							E2403	No Clinical Signs, Symptoms or Conditions	A051201	Device Dislodged or Dislocated	05/10/2022	05/10/2022
									A24	No Apparent Device Problem	05/10/2022	05/10/2022
							F1203	Life Threatening Illness or Injury	A051201	Device Dislodged or Dislocated	05/10/2022	05/10/2022
									A24	No Apparent Device Problem	05/10/2022	05/10/2022
							F2303	Medication Required	A051201	Device Dislodged or Dislocated	05/10/2022	05/10/2022
									A24	No Apparent Device Problem	05/10/2022	05/10/2022
8050	2022	All other reportable incidents	35223	Endocardial/interventricular septal pacing lead	CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A27	Appropriate Device Problem Term/Code not Available		
							F23	Unexpected Medical Intervention	A27	Appropriate Device Problem Term/Code not Available		
8051	2022	All other reportable incidents	35223	Endocardial/interventricular septal pacing lead	CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A040101	Fracture		
									A070909	Oversensing		
							F1901	Additional Surgery	A040101	Fracture		
									A070909	Oversensing		
8052	2022	All other reportable incidents	35223	Endocardial/interventricular septal pacing lead	CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A090801	Signal Artifact/Noise		
							F26	No Health Consequences or Impact	A090801	Signal Artifact/Noise		
8053	2022	All other reportable incidents	47265	Dual-chamber implantable pacemaker, rate-responsive	CT141	Electrodes/Leads and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	31/10/2022	31/10/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	31/10/2022	31/10/2022
					CT145	Cardiovascular devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	31/10/2022	31/10/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	31/10/2022	31/10/2022
					CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	31/10/2022	31/10/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	31/10/2022	31/10/2022
					CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	31/10/2022	31/10/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	31/10/2022	31/10/2022
					CT233	Surgical	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	31/10/2022	31/10/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	31/10/2022	31/10/2022
					CT2406	Implantable	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	31/10/2022	31/10/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	31/10/2022	31/10/2022
					CT2407	Active implantable	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	31/10/2022	31/10/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	31/10/2022	31/10/2022
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	31/10/2022	31/10/2022

						F1901	Additional Surgery	A24	No Apparent Device Problem	31/10/2022	31/10/2022
				CT2727	Active implant devices/ancillaries	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	31/10/2022	31/10/2022
						F1901	Additional Surgery	A24	No Apparent Device Problem	31/10/2022	31/10/2022
				CT2750	Implantable cardiac electrical stimulators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	31/10/2022	31/10/2022
						F1901	Additional Surgery	A24	No Apparent Device Problem	31/10/2022	31/10/2022
				CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	31/10/2022	31/10/2022
						F1901	Additional Surgery	A24	No Apparent Device Problem	31/10/2022	31/10/2022
				CT2985	Active	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	31/10/2022	31/10/2022
						F1901	Additional Surgery	A24	No Apparent Device Problem	31/10/2022	31/10/2022
				CT3001	Implantable metal	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	31/10/2022	31/10/2022
						F1901	Additional Surgery	A24	No Apparent Device Problem	31/10/2022	31/10/2022
				CT3002	Implantable synthetic polymer	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	31/10/2022	31/10/2022
						F1901	Additional Surgery	A24	No Apparent Device Problem	31/10/2022	31/10/2022
				CT3006	Implantable inorganic material	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	31/10/2022	31/10/2022
						F1901	Additional Surgery	A24	No Apparent Device Problem	31/10/2022	31/10/2022
				CT314	Electrically powered	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	31/10/2022	31/10/2022
						F1901	Additional Surgery	A24	No Apparent Device Problem	31/10/2022	31/10/2022
				CT321	Long-term surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	31/10/2022	31/10/2022
						F1901	Additional Surgery	A24	No Apparent Device Problem	31/10/2022	31/10/2022
				CT449	Generators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	31/10/2022	31/10/2022
						F1901	Additional Surgery	A24	No Apparent Device Problem	31/10/2022	31/10/2022
				CT981	Single use	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	31/10/2022	31/10/2022
						F1901	Additional Surgery	A24	No Apparent Device Problem	31/10/2022	31/10/2022
				CT983	Surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	31/10/2022	31/10/2022
						F1901	Additional Surgery	A24	No Apparent Device Problem	31/10/2022	31/10/2022
8054	2022	All other reportable incidents	35223	Endocardial/interventricular septal pacing lead	CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070909	Oversensing		
							F26	No Health Consequences or Impact	A070909	Oversensing		
8055	2022	Death	17882	Physiologic-monitoring defibrillation system	CT1684	Cardiac pacemakers and associated devices	E0602	Cardiac Arrest	A071301	Failure to Deliver Shock/Stimulation		
							F02	Death	A071301	Failure to Deliver Shock/Stimulation		
8056	2022	All other reportable incidents	17882	Physiologic-monitoring defibrillation system	CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A07	Electrical /Electronic Property Problem		
							F26	No Health Consequences or Impact	A07	Electrical /Electronic Property Problem		
8057	2021	All other reportable incidents	36241	Pacemaker/defibrillator lead	CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A05	Mechanical Problem	24/08/2021	24/08/2021
									A070102	High Capture Threshold	24/08/2021	24/08/2021
									A070910	Undersensing	24/08/2021	24/08/2021
									A072201	High Impedance	24/08/2021	24/08/2021
									A1301	Failure to Read Input Signal	24/08/2021	24/08/2021
									A2304	Off-Label Use	24/08/2021	24/08/2021
							F1908	Prolonged Surgery	A05	Mechanical Problem	24/08/2021	24/08/2021
									A070102	High Capture Threshold	24/08/2021	24/08/2021
									A070910	Undersensing	24/08/2021	24/08/2021
									A072201	High Impedance	24/08/2021	24/08/2021
									A1301	Failure to Read Input Signal	24/08/2021	24/08/2021
									A2304	Off-Label Use	24/08/2021	24/08/2021
8058	2021	All other reportable incidents	47265	Dual-chamber implantable pacemaker, rate-responsive	CT141	Electrodes/Leads and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem		
									A072201	High Impedance		
							F26	No Health Consequences or Impact	A0705	Battery Problem		
									A072201	High Impedance		
					CT145	Cardiovascular devices	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem		
									A072201	High Impedance		
							F26	No Health Consequences or Impact	A0705	Battery Problem		
									A072201	High Impedance		
					CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem		
									A072201	High Impedance		
							F26	No Health Consequences or Impact	A0705	Battery Problem		
									A072201	High Impedance		
					CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem		
									A072201	High Impedance		
							F26	No Health Consequences or Impact	A0705	Battery Problem		
									A072201	High Impedance		
					CT233	Surgical	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem		
									A072201	High Impedance		
							F26	No Health Consequences or Impact	A0705	Battery Problem		
									A072201	High Impedance		
					CT2406	Implantable	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem		
									A072201	High Impedance		
							F26	No Health Consequences or Impact	A0705	Battery Problem		
									A072201	High Impedance		
					CT2407	Active implantable	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem		
									A072201	High Impedance		
							F26	No Health Consequences or Impact	A0705	Battery Problem		
									A072201	High Impedance		
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem		
									A072201	High Impedance		
							F26	No Health Consequences or Impact	A0705	Battery Problem		
									A072201	High Impedance		
					CT2727	Active implant devices/ancillaries	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem		
									A072201	High Impedance		
							F26	No Health Consequences or Impact	A0705	Battery Problem		
									A072201	High Impedance		
					CT2750	Implantable cardiac electrical stimulators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem		
									A072201	High Impedance		
							F26	No Health Consequences or Impact	A0705	Battery Problem		
									A072201	High Impedance		
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem		
									A072201	High Impedance		
							F26	No Health Consequences or Impact	A0705	Battery Problem		
									A072201	High Impedance		
					CT2985	Active	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem		

									A072201	High Impedance		
							F26	No Health Consequences or Impact	A0705	Battery Problem		
									A072201	High Impedance		
					CT3001	Implantable metal	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem		
									A072201	High Impedance		
							F26	No Health Consequences or Impact	A0705	Battery Problem		
									A072201	High Impedance		
					CT3002	Implantable synthetic polymer	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem		
									A072201	High Impedance		
							F26	No Health Consequences or Impact	A0705	Battery Problem		
									A072201	High Impedance		
					CT3006	Implantable inorganic material	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem		
									A072201	High Impedance		
							F26	No Health Consequences or Impact	A0705	Battery Problem		
									A072201	High Impedance		
					CT314	Electrically powered	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem		
									A072201	High Impedance		
							F26	No Health Consequences or Impact	A0705	Battery Problem		
									A072201	High Impedance		
					CT321	Long-term surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem		
									A072201	High Impedance		
							F26	No Health Consequences or Impact	A0705	Battery Problem		
									A072201	High Impedance		
					CT449	Generators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem		
									A072201	High Impedance		
							F26	No Health Consequences or Impact	A0705	Battery Problem		
									A072201	High Impedance		
					CT981	Single use	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem		
									A072201	High Impedance		
							F26	No Health Consequences or Impact	A0705	Battery Problem		
									A072201	High Impedance		
					CT983	Surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem		
									A072201	High Impedance		
							F26	No Health Consequences or Impact	A0705	Battery Problem		
									A072201	High Impedance		
8059	2022	All other reportable incidents	60789	Intracardiac pacemaker	CT130	Catheters and associated devices	E0119	Loss of Consciousness	A23	Use of Device Problem		
							E0604	Cardiac Perforation	A23	Use of Device Problem		
							E0605	Cardiac Tamponade	A23	Use of Device Problem		
							E0619	Pericardial Effusion	A23	Use of Device Problem		
							E2403	No Clinical Signs, Symptoms or Conditions	A23	Use of Device Problem		
							F08	Hospitalization or Prolonged Hospitalization	A23	Use of Device Problem		
							F1203	Life Threatening Illness or Injury	A23	Use of Device Problem		
					CT145	Cardiovascular devices	E0119	Loss of Consciousness	A23	Use of Device Problem		
							E0604	Cardiac Perforation	A23	Use of Device Problem		
							E0605	Cardiac Tamponade	A23	Use of Device Problem		
							E0619	Pericardial Effusion	A23	Use of Device Problem		
							E2403	No Clinical Signs, Symptoms or Conditions	A23	Use of Device Problem		
							F08	Hospitalization or Prolonged Hospitalization	A23	Use of Device Problem		
							F1203	Life Threatening Illness or Injury	A23	Use of Device Problem		
					CT1684	Cardiac pacemakers and associated devices	E0119	Loss of Consciousness	A23	Use of Device Problem		
							E0604	Cardiac Perforation	A23	Use of Device Problem		
							E0605	Cardiac Tamponade	A23	Use of Device Problem		
							E0619	Pericardial Effusion	A23	Use of Device Problem		
							E2403	No Clinical Signs, Symptoms or Conditions	A23	Use of Device Problem		
							F08	Hospitalization or Prolonged Hospitalization	A23	Use of Device Problem		
							F1203	Life Threatening Illness or Injury	A23	Use of Device Problem		
					CT1685	Cardiac pacemakers	E0119	Loss of Consciousness	A23	Use of Device Problem		
							E0604	Cardiac Perforation	A23	Use of Device Problem		
							E0605	Cardiac Tamponade	A23	Use of Device Problem		
							E0619	Pericardial Effusion	A23	Use of Device Problem		
							E2403	No Clinical Signs, Symptoms or Conditions	A23	Use of Device Problem		
							F08	Hospitalization or Prolonged Hospitalization	A23	Use of Device Problem		
							F1203	Life Threatening Illness or Injury	A23	Use of Device Problem		
					CT233	Surgical	E0119	Loss of Consciousness	A23	Use of Device Problem		
							E0604	Cardiac Perforation	A23	Use of Device Problem		
							E0605	Cardiac Tamponade	A23	Use of Device Problem		
							E0619	Pericardial Effusion	A23	Use of Device Problem		
							E2403	No Clinical Signs, Symptoms or Conditions	A23	Use of Device Problem		
							F08	Hospitalization or Prolonged Hospitalization	A23	Use of Device Problem		
							F1203	Life Threatening Illness or Injury	A23	Use of Device Problem		
					CT2406	Implantable	E0119	Loss of Consciousness	A23	Use of Device Problem		
							E0604	Cardiac Perforation	A23	Use of Device Problem		
							E0605	Cardiac Tamponade	A23	Use of Device Problem		
							E0619	Pericardial Effusion	A23	Use of Device Problem		
							E2403	No Clinical Signs, Symptoms or Conditions	A23	Use of Device Problem		
							F08	Hospitalization or Prolonged Hospitalization	A23	Use of Device Problem		
							F1203	Life Threatening Illness or Injury	A23	Use of Device Problem		
					CT2407	Active implantable	E0119	Loss of Consciousness	A23	Use of Device Problem		
							E0604	Cardiac Perforation	A23	Use of Device Problem		
							E0605	Cardiac Tamponade	A23	Use of Device Problem		
							E0619	Pericardial Effusion	A23	Use of Device Problem		
							E2403	No Clinical Signs, Symptoms or Conditions	A23	Use of Device Problem		
							F08	Hospitalization or Prolonged Hospitalization	A23	Use of Device Problem		
							F1203	Life Threatening Illness or Injury	A23	Use of Device Problem		
					CT257	Implantable cardiac pacemakers	E0119	Loss of Consciousness	A23	Use of Device Problem		
							E0604	Cardiac Perforation	A23	Use of Device Problem		
							E0605	Cardiac Tamponade	A23	Use of Device Problem		
							E0619	Pericardial Effusion	A23	Use of Device Problem		

							E2403	No Clinical Signs, Symptoms or Conditions	A23	Use of Device Problem		
							F08	Hospitalization or Prolonged Hospitalization	A23	Use of Device Problem		
							F1203	Life Threatening Illness or Injury	A23	Use of Device Problem		
					CT2727	Active implant devices/ancillaries	E0119	Loss of Consciousness	A23	Use of Device Problem		
							E0604	Cardiac Perforation	A23	Use of Device Problem		
							E0605	Cardiac Tamponade	A23	Use of Device Problem		
							E0619	Pericardial Effusion	A23	Use of Device Problem		
							E2403	No Clinical Signs, Symptoms or Conditions	A23	Use of Device Problem		
							F08	Hospitalization or Prolonged Hospitalization	A23	Use of Device Problem		
							F1203	Life Threatening Illness or Injury	A23	Use of Device Problem		
					CT2750	Implantable cardiac electrical stimulators and associated devices	E0119	Loss of Consciousness	A23	Use of Device Problem		
							E0604	Cardiac Perforation	A23	Use of Device Problem		
							E0605	Cardiac Tamponade	A23	Use of Device Problem		
							E0619	Pericardial Effusion	A23	Use of Device Problem		
							E2403	No Clinical Signs, Symptoms or Conditions	A23	Use of Device Problem		
							F08	Hospitalization or Prolonged Hospitalization	A23	Use of Device Problem		
							F1203	Life Threatening Illness or Injury	A23	Use of Device Problem		
					CT2818	Catheterization/Cannulation/Fluid path devices	E0119	Loss of Consciousness	A23	Use of Device Problem		
							E0604	Cardiac Perforation	A23	Use of Device Problem		
							E0605	Cardiac Tamponade	A23	Use of Device Problem		
							E0619	Pericardial Effusion	A23	Use of Device Problem		
							E2403	No Clinical Signs, Symptoms or Conditions	A23	Use of Device Problem		
							F08	Hospitalization or Prolonged Hospitalization	A23	Use of Device Problem		
							F1203	Life Threatening Illness or Injury	A23	Use of Device Problem		
					CT2866	Implantable cardiac electrical stimulators	E0119	Loss of Consciousness	A23	Use of Device Problem		
							E0604	Cardiac Perforation	A23	Use of Device Problem		
							E0605	Cardiac Tamponade	A23	Use of Device Problem		
							E0619	Pericardial Effusion	A23	Use of Device Problem		
							E2403	No Clinical Signs, Symptoms or Conditions	A23	Use of Device Problem		
							F08	Hospitalization or Prolonged Hospitalization	A23	Use of Device Problem		
							F1203	Life Threatening Illness or Injury	A23	Use of Device Problem		
					CT2985	Active	E0119	Loss of Consciousness	A23	Use of Device Problem		
							E0604	Cardiac Perforation	A23	Use of Device Problem		
							E0605	Cardiac Tamponade	A23	Use of Device Problem		
							E0619	Pericardial Effusion	A23	Use of Device Problem		
							E2403	No Clinical Signs, Symptoms or Conditions	A23	Use of Device Problem		
							F08	Hospitalization or Prolonged Hospitalization	A23	Use of Device Problem		
							F1203	Life Threatening Illness or Injury	A23	Use of Device Problem		
					CT3001	Implantable metal	E0119	Loss of Consciousness	A23	Use of Device Problem		
							E0604	Cardiac Perforation	A23	Use of Device Problem		
							E0605	Cardiac Tamponade	A23	Use of Device Problem		
							E0619	Pericardial Effusion	A23	Use of Device Problem		
							E2403	No Clinical Signs, Symptoms or Conditions	A23	Use of Device Problem		
							F08	Hospitalization or Prolonged Hospitalization	A23	Use of Device Problem		
							F1203	Life Threatening Illness or Injury	A23	Use of Device Problem		
					CT3002	Implantable synthetic polymer	E0119	Loss of Consciousness	A23	Use of Device Problem		
							E0604	Cardiac Perforation	A23	Use of Device Problem		
							E0605	Cardiac Tamponade	A23	Use of Device Problem		
							E0619	Pericardial Effusion	A23	Use of Device Problem		
							E2403	No Clinical Signs, Symptoms or Conditions	A23	Use of Device Problem		
							F08	Hospitalization or Prolonged Hospitalization	A23	Use of Device Problem		
							F1203	Life Threatening Illness or Injury	A23	Use of Device Problem		
					CT3006	Implantable inorganic material	E0119	Loss of Consciousness	A23	Use of Device Problem		
							E0604	Cardiac Perforation	A23	Use of Device Problem		
							E0605	Cardiac Tamponade	A23	Use of Device Problem		
							E0619	Pericardial Effusion	A23	Use of Device Problem		
							E2403	No Clinical Signs, Symptoms or Conditions	A23	Use of Device Problem		
							F08	Hospitalization or Prolonged Hospitalization	A23	Use of Device Problem		
							F1203	Life Threatening Illness or Injury	A23	Use of Device Problem		
					CT314	Electrically powered	E0119	Loss of Consciousness	A23	Use of Device Problem		
							E0604	Cardiac Perforation	A23	Use of Device Problem		
							E0605	Cardiac Tamponade	A23	Use of Device Problem		
							E0619	Pericardial Effusion	A23	Use of Device Problem		
							E2403	No Clinical Signs, Symptoms or Conditions	A23	Use of Device Problem		
							F08	Hospitalization or Prolonged Hospitalization	A23	Use of Device Problem		
							F1203	Life Threatening Illness or Injury	A23	Use of Device Problem		
					CT321	Long-term surgical invasive	E0119	Loss of Consciousness	A23	Use of Device Problem		
							E0604	Cardiac Perforation	A23	Use of Device Problem		
							E0605	Cardiac Tamponade	A23	Use of Device Problem		
							E0619	Pericardial Effusion	A23	Use of Device Problem		
							E2403	No Clinical Signs, Symptoms or Conditions	A23	Use of Device Problem		
							F08	Hospitalization or Prolonged Hospitalization	A23	Use of Device Problem		
							F1203	Life Threatening Illness or Injury	A23	Use of Device Problem		
					CT981	Single use	E0119	Loss of Consciousness	A23	Use of Device Problem		
							E0604	Cardiac Perforation	A23	Use of Device Problem		
							E0605	Cardiac Tamponade	A23	Use of Device Problem		
							E0619	Pericardial Effusion	A23	Use of Device Problem		
							E2403	No Clinical Signs, Symptoms or Conditions	A23	Use of Device Problem		
							F08	Hospitalization or Prolonged Hospitalization	A23	Use of Device Problem		
							F1203	Life Threatening Illness or Injury	A23	Use of Device Problem		
					CT983	Surgical invasive	E0119	Loss of Consciousness	A23	Use of Device Problem		
							E0604	Cardiac Perforation	A23	Use of Device Problem		
							E0605	Cardiac Tamponade	A23	Use of Device Problem		
							E0619	Pericardial Effusion	A23	Use of Device Problem		
							E2403	No Clinical Signs, Symptoms or Conditions	A23	Use of Device Problem		
							F08	Hospitalization or Prolonged Hospitalization	A23	Use of Device Problem		
							F1203	Life Threatening Illness or Injury	A23	Use of Device Problem		
8060	2022	All other reportable incidents	35223	Endocardial/interventricular septal pacing lead	CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A090801	Signal Artifact/Noise		

							F12	Serious Injury/ Illness/ Impairment	A090801	Signal Artifact/Noise		
							F23	Unexpected Medical Intervention	A090801	Signal Artifact/Noise		
8061	2022	All other reportable incidents	35223	Endocardial/interventricular septal pacing lead	CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070901	Decreased Sensitivity		
									A070909	Oversensing		
									A070910	Undersensing		
									A0712	Pacing Problem		
									A072202	Low Impedance		
							F23	Unexpected Medical Intervention	A070901	Decreased Sensitivity		
									A070909	Oversensing		
									A070910	Undersensing		
									A0712	Pacing Problem		
									A072202	Low Impedance		
8062	2022	All other reportable incidents	35223	Endocardial/interventricular septal pacing lead	CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070909	Oversensing		
									A0712	Pacing Problem		
									A090801	Signal Artifact/Noise		
							F26	No Health Consequences or Impact	A070909	Oversensing		
									A0712	Pacing Problem		
									A090801	Signal Artifact/Noise		
8063	2021	All other reportable incidents	47265	Dual-chamber implantable pacemaker, rate-responsive	CT141	Electrodes/Leads and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	20/10/2021	20/10/2021
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	20/10/2021	20/10/2021
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	20/10/2021	20/10/2021
					CT145	Cardiovascular devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	20/10/2021	20/10/2021
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	20/10/2021	20/10/2021
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	20/10/2021	20/10/2021
					CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	20/10/2021	20/10/2021
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	20/10/2021	20/10/2021
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	20/10/2021	20/10/2021
					CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	20/10/2021	20/10/2021
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	20/10/2021	20/10/2021
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	20/10/2021	20/10/2021
					CT233	Surgical	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	20/10/2021	20/10/2021
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	20/10/2021	20/10/2021
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	20/10/2021	20/10/2021
					CT2406	Implantable	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	20/10/2021	20/10/2021
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	20/10/2021	20/10/2021
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	20/10/2021	20/10/2021
					CT2407	Active implantable	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	20/10/2021	20/10/2021
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	20/10/2021	20/10/2021
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	20/10/2021	20/10/2021
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	20/10/2021	20/10/2021
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	20/10/2021	20/10/2021
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	20/10/2021	20/10/2021
					CT2727	Active implant devices/ancillaries	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	20/10/2021	20/10/2021
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	20/10/2021	20/10/2021
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	20/10/2021	20/10/2021
					CT2750	Implantable cardiac electrical stimulators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	20/10/2021	20/10/2021
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	20/10/2021	20/10/2021
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	20/10/2021	20/10/2021
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	20/10/2021	20/10/2021
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	20/10/2021	20/10/2021
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	20/10/2021	20/10/2021
					CT2985	Active	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	20/10/2021	20/10/2021
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	20/10/2021	20/10/2021
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	20/10/2021	20/10/2021
					CT3001	Implantable metal	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	20/10/2021	20/10/2021
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	20/10/2021	20/10/2021
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	20/10/2021	20/10/2021
					CT3002	Implantable synthetic polymer	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	20/10/2021	20/10/2021
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	20/10/2021	20/10/2021
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	20/10/2021	20/10/2021
					CT3006	Implantable inorganic material	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	20/10/2021	20/10/2021
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	20/10/2021	20/10/2021
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	20/10/2021	20/10/2021
					CT314	Electrically powered	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	20/10/2021	20/10/2021
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	20/10/2021	20/10/2021
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	20/10/2021	20/10/2021
					CT321	Long-term surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	20/10/2021	20/10/2021
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	20/10/2021	20/10/2021
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	20/10/2021	20/10/2021
					CT449	Generators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	20/10/2021	20/10/2021
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	20/10/2021	20/10/2021
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	20/10/2021	20/10/2021
					CT981	Single use	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	20/10/2021	20/10/2021
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	20/10/2021	20/10/2021
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	20/10/2021	20/10/2021
					CT983	Surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	20/10/2021	20/10/2021
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	20/10/2021	20/10/2021
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	20/10/2021	20/10/2021
8064	2022	All other reportable incidents	35223	Endocardial/interventricular septal pacing lead	CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A040101	Fracture		
									A072201	High Impedance		
							F26	No Health Consequences or Impact	A040101	Fracture		
									A072201	High Impedance		
8065	2022	All other reportable incidents	47265	Dual-chamber implantable pacemaker, rate-responsive	CT141	Electrodes/Leads and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A0908	Incorrect, Inadequate or Imprecise Result or Readings	08/02/2022	08/02/2022
							F26	No Health Consequences or Impact	A0908	Incorrect, Inadequate or Imprecise Result or Readings	08/02/2022	08/02/2022
					CT145	Cardiovascular devices	E2403	No Clinical Signs, Symptoms or Conditions	A0908	Incorrect, Inadequate or Imprecise Result or Readings	08/02/2022	08/02/2022

							F26	No Health Consequences or Impact	A0908	Incorrect, Inadequate or Imprecise Result or Readings	08/02/2022	08/02/2022
					CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A0908	Incorrect, Inadequate or Imprecise Result or Readings	08/02/2022	08/02/2022
							F26	No Health Consequences or Impact	A0908	Incorrect, Inadequate or Imprecise Result or Readings	08/02/2022	08/02/2022
					CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A0908	Incorrect, Inadequate or Imprecise Result or Readings	08/02/2022	08/02/2022
							F26	No Health Consequences or Impact	A0908	Incorrect, Inadequate or Imprecise Result or Readings	08/02/2022	08/02/2022
					CT233	Surgical	E2403	No Clinical Signs, Symptoms or Conditions	A0908	Incorrect, Inadequate or Imprecise Result or Readings	08/02/2022	08/02/2022
							F26	No Health Consequences or Impact	A0908	Incorrect, Inadequate or Imprecise Result or Readings	08/02/2022	08/02/2022
					CT2406	Implantable	E2403	No Clinical Signs, Symptoms or Conditions	A0908	Incorrect, Inadequate or Imprecise Result or Readings	08/02/2022	08/02/2022
							F26	No Health Consequences or Impact	A0908	Incorrect, Inadequate or Imprecise Result or Readings	08/02/2022	08/02/2022
					CT2407	Active implantable	E2403	No Clinical Signs, Symptoms or Conditions	A0908	Incorrect, Inadequate or Imprecise Result or Readings	08/02/2022	08/02/2022
							F26	No Health Consequences or Impact	A0908	Incorrect, Inadequate or Imprecise Result or Readings	08/02/2022	08/02/2022
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A0908	Incorrect, Inadequate or Imprecise Result or Readings	08/02/2022	08/02/2022
							F26	No Health Consequences or Impact	A0908	Incorrect, Inadequate or Imprecise Result or Readings	08/02/2022	08/02/2022
					CT2727	Active implant devices/ancillaries	E2403	No Clinical Signs, Symptoms or Conditions	A0908	Incorrect, Inadequate or Imprecise Result or Readings	08/02/2022	08/02/2022
							F26	No Health Consequences or Impact	A0908	Incorrect, Inadequate or Imprecise Result or Readings	08/02/2022	08/02/2022
					CT2750	Implantable cardiac electrical stimulators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A0908	Incorrect, Inadequate or Imprecise Result or Readings	08/02/2022	08/02/2022
							F26	No Health Consequences or Impact	A0908	Incorrect, Inadequate or Imprecise Result or Readings	08/02/2022	08/02/2022
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A0908	Incorrect, Inadequate or Imprecise Result or Readings	08/02/2022	08/02/2022
							F26	No Health Consequences or Impact	A0908	Incorrect, Inadequate or Imprecise Result or Readings	08/02/2022	08/02/2022
					CT2985	Active	E2403	No Clinical Signs, Symptoms or Conditions	A0908	Incorrect, Inadequate or Imprecise Result or Readings	08/02/2022	08/02/2022
							F26	No Health Consequences or Impact	A0908	Incorrect, Inadequate or Imprecise Result or Readings	08/02/2022	08/02/2022
					CT3001	Implantable metal	E2403	No Clinical Signs, Symptoms or Conditions	A0908	Incorrect, Inadequate or Imprecise Result or Readings	08/02/2022	08/02/2022
							F26	No Health Consequences or Impact	A0908	Incorrect, Inadequate or Imprecise Result or Readings	08/02/2022	08/02/2022
					CT3002	Implantable synthetic polymer	E2403	No Clinical Signs, Symptoms or Conditions	A0908	Incorrect, Inadequate or Imprecise Result or Readings	08/02/2022	08/02/2022
							F26	No Health Consequences or Impact	A0908	Incorrect, Inadequate or Imprecise Result or Readings	08/02/2022	08/02/2022
					CT3006	Implantable inorganic material	E2403	No Clinical Signs, Symptoms or Conditions	A0908	Incorrect, Inadequate or Imprecise Result or Readings	08/02/2022	08/02/2022
							F26	No Health Consequences or Impact	A0908	Incorrect, Inadequate or Imprecise Result or Readings	08/02/2022	08/02/2022
					CT314	Electrically powered	E2403	No Clinical Signs, Symptoms or Conditions	A0908	Incorrect, Inadequate or Imprecise Result or Readings	08/02/2022	08/02/2022
							F26	No Health Consequences or Impact	A0908	Incorrect, Inadequate or Imprecise Result or Readings	08/02/2022	08/02/2022
					CT321	Long-term surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A0908	Incorrect, Inadequate or Imprecise Result or Readings	08/02/2022	08/02/2022
							F26	No Health Consequences or Impact	A0908	Incorrect, Inadequate or Imprecise Result or Readings	08/02/2022	08/02/2022
					CT449	Generators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A0908	Incorrect, Inadequate or Imprecise Result or Readings	08/02/2022	08/02/2022
							F26	No Health Consequences or Impact	A0908	Incorrect, Inadequate or Imprecise Result or Readings	08/02/2022	08/02/2022
					CT981	Single use	E2403	No Clinical Signs, Symptoms or Conditions	A0908	Incorrect, Inadequate or Imprecise Result or Readings	08/02/2022	08/02/2022
							F26	No Health Consequences or Impact	A0908	Incorrect, Inadequate or Imprecise Result or Readings	08/02/2022	08/02/2022
					CT983	Surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A0908	Incorrect, Inadequate or Imprecise Result or Readings	08/02/2022	08/02/2022
							F26	No Health Consequences or Impact	A0908	Incorrect, Inadequate or Imprecise Result or Readings	08/02/2022	08/02/2022
8066	2022	All other reportable incidents	47265	Dual-chamber implantable pacemaker, rate-responsive	CT141	Electrodes/Leads and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070102	High Capture Threshold		
							F26	No Health Consequences or Impact	A070102	High Capture Threshold		
					CT145	Cardiovascular devices	E2403	No Clinical Signs, Symptoms or Conditions	A070102	High Capture Threshold		
							F26	No Health Consequences or Impact	A070102	High Capture Threshold		
					CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070102	High Capture Threshold		
							F26	No Health Consequences or Impact	A070102	High Capture Threshold		
					CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A070102	High Capture Threshold		
							F26	No Health Consequences or Impact	A070102	High Capture Threshold		
					CT233	Surgical	E2403	No Clinical Signs, Symptoms or Conditions	A070102	High Capture Threshold		
							F26	No Health Consequences or Impact	A070102	High Capture Threshold		
					CT2406	Implantable	E2403	No Clinical Signs, Symptoms or Conditions	A070102	High Capture Threshold		
							F26	No Health Consequences or Impact	A070102	High Capture Threshold		
					CT2407	Active implantable	E2403	No Clinical Signs, Symptoms or Conditions	A070102	High Capture Threshold		
							F26	No Health Consequences or Impact	A070102	High Capture Threshold		
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A070102	High Capture Threshold		
							F26	No Health Consequences or Impact	A070102	High Capture Threshold		
					CT2727	Active implant devices/ancillaries	E2403	No Clinical Signs, Symptoms or Conditions	A070102	High Capture Threshold		
							F26	No Health Consequences or Impact	A070102	High Capture Threshold		
					CT2750	Implantable cardiac electrical stimulators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070102	High Capture Threshold		
							F26	No Health Consequences or Impact	A070102	High Capture Threshold		
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A070102	High Capture Threshold		
							F26	No Health Consequences or Impact	A070102	High Capture Threshold		
					CT2985	Active	E2403	No Clinical Signs, Symptoms or Conditions	A070102	High Capture Threshold		

						F26	No Health Consequences or Impact	A070102	High Capture Threshold		
				CT3001	Implantable metal	E2403	No Clinical Signs, Symptoms or Conditions	A070102	High Capture Threshold		
						F26	No Health Consequences or Impact	A070102	High Capture Threshold		
				CT3002	Implantable synthetic polymer	E2403	No Clinical Signs, Symptoms or Conditions	A070102	High Capture Threshold		
						F26	No Health Consequences or Impact	A070102	High Capture Threshold		
				CT3006	Implantable inorganic material	E2403	No Clinical Signs, Symptoms or Conditions	A070102	High Capture Threshold		
						F26	No Health Consequences or Impact	A070102	High Capture Threshold		
				CT314	Electrically powered	E2403	No Clinical Signs, Symptoms or Conditions	A070102	High Capture Threshold		
						F26	No Health Consequences or Impact	A070102	High Capture Threshold		
				CT321	Long-term surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A070102	High Capture Threshold		
						F26	No Health Consequences or Impact	A070102	High Capture Threshold		
				CT449	Generators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070102	High Capture Threshold		
						F26	No Health Consequences or Impact	A070102	High Capture Threshold		
				CT981	Single use	E2403	No Clinical Signs, Symptoms or Conditions	A070102	High Capture Threshold		
						F26	No Health Consequences or Impact	A070102	High Capture Threshold		
				CT983	Surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A070102	High Capture Threshold		
						F26	No Health Consequences or Impact	A070102	High Capture Threshold		
8067	2022	All other reportable incidents	35223	Endocardial/interventricular septal pacing lead	CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070901	Decreased Sensitivity		
									A070909	Oversensing		
									A070910	Undersensing		
							F23	Unexpected Medical Intervention	A070901	Decreased Sensitivity		
									A070909	Oversensing		
									A070910	Undersensing		
							F26	No Health Consequences or Impact	A070901	Decreased Sensitivity		
									A070909	Oversensing		
									A070910	Undersensing		
8068	2022		35852	Single-chamber implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E0601	Arrhythmia	A070602	Delayed Charge Time	11/11/2022	11/11/2022
									A071302	Inappropriate/Inadequate Shock/Stimulation	11/11/2022	11/11/2022
									A160503	Inappropriate or Unexpected Reset	11/11/2022	11/11/2022
							F02	Death	A070602	Delayed Charge Time	11/11/2022	11/11/2022
									A071302	Inappropriate/Inadequate Shock/Stimulation	11/11/2022	11/11/2022
									A160503	Inappropriate or Unexpected Reset	11/11/2022	11/11/2022
8069	2022		47265	Dual-chamber implantable pacemaker, rate-responsive	CT141	Electrodes/Leads and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem		
							F02	Death	A24	No Apparent Device Problem		
					CT145	Cardiovascular devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem		
							F02	Death	A24	No Apparent Device Problem		
					CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem		
							F02	Death	A24	No Apparent Device Problem		
					CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem		
							F02	Death	A24	No Apparent Device Problem		
					CT233	Surgical	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem		
							F02	Death	A24	No Apparent Device Problem		
					CT2406	Implantable	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem		
							F02	Death	A24	No Apparent Device Problem		
					CT2407	Active implantable	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem		
							F02	Death	A24	No Apparent Device Problem		
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem		
							F02	Death	A24	No Apparent Device Problem		
					CT2727	Active implant devices/ancillaries	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem		
							F02	Death	A24	No Apparent Device Problem		
					CT2750	Implantable cardiac electrical stimulators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem		
							F02	Death	A24	No Apparent Device Problem		
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem		
							F02	Death	A24	No Apparent Device Problem		
					CT2985	Active	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem		
							F02	Death	A24	No Apparent Device Problem		
					CT3001	Implantable metal	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem		
							F02	Death	A24	No Apparent Device Problem		
					CT3002	Implantable synthetic polymer	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem		
							F02	Death	A24	No Apparent Device Problem		
					CT3006	Implantable inorganic material	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem		
							F02	Death	A24	No Apparent Device Problem		
					CT314	Electrically powered	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem		
							F02	Death	A24	No Apparent Device Problem		
					CT321	Long-term surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem		
							F02	Death	A24	No Apparent Device Problem		
					CT449	Generators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem		
							F02	Death	A24	No Apparent Device Problem		
					CT981	Single use	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem		
							F02	Death	A24	No Apparent Device Problem		
					CT983	Surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem		
							F02	Death	A24	No Apparent Device Problem		
8070	2022	Death	47265	Dual-chamber implantable pacemaker, rate-responsive	CT141	Electrodes/Leads and associated devices	E060109	Tachycardia	A09	Output Problem		
									A090811	Unable to Obtain Readings		
							E060110	Ventricular Fibrillation	A09	Output Problem		
									A090811	Unable to Obtain Readings		
							F02	Death	A09	Output Problem		
									A090811	Unable to Obtain Readings		
					CT145	Cardiovascular devices	E060109	Tachycardia	A09	Output Problem		
									A090811	Unable to Obtain Readings		
							E060110	Ventricular Fibrillation	A09	Output Problem		
									A090811	Unable to Obtain Readings		
							F02	Death	A09	Output Problem		
									A090811	Unable to Obtain Readings		
					CT1684	Cardiac pacemakers and associated devices	E060109	Tachycardia	A09	Output Problem		
									A090811	Unable to Obtain Readings		
							E060110	Ventricular Fibrillation	A09	Output Problem		
									A090811	Unable to Obtain Readings		

						F02	Death	A09	Output Problem		
								A090811	Unable to Obtain Readings		
					CT1685	Cardiac pacemakers	E060109	Tachycardia	A09	Output Problem		
									A090811	Unable to Obtain Readings		
							E060110	Ventricular Fibrillation	A09	Output Problem		
									A090811	Unable to Obtain Readings		
							F02	Death	A09	Output Problem		
									A090811	Unable to Obtain Readings		
					CT233	Surgical	E060109	Tachycardia	A09	Output Problem		
									A090811	Unable to Obtain Readings		
							E060110	Ventricular Fibrillation	A09	Output Problem		
									A090811	Unable to Obtain Readings		
							F02	Death	A09	Output Problem		
									A090811	Unable to Obtain Readings		
					CT2406	Implantable	E060109	Tachycardia	A09	Output Problem		
									A090811	Unable to Obtain Readings		
							E060110	Ventricular Fibrillation	A09	Output Problem		
									A090811	Unable to Obtain Readings		
							F02	Death	A09	Output Problem		
									A090811	Unable to Obtain Readings		
					CT2407	Active implantable	E060109	Tachycardia	A09	Output Problem		
									A090811	Unable to Obtain Readings		
							E060110	Ventricular Fibrillation	A09	Output Problem		
									A090811	Unable to Obtain Readings		
							F02	Death	A09	Output Problem		
									A090811	Unable to Obtain Readings		
					CT257	Implantable cardiac pacemakers	E060109	Tachycardia	A09	Output Problem		
									A090811	Unable to Obtain Readings		
							E060110	Ventricular Fibrillation	A09	Output Problem		
									A090811	Unable to Obtain Readings		
							F02	Death	A09	Output Problem		
									A090811	Unable to Obtain Readings		
					CT2727	Active implant devices/ancillaries	E060109	Tachycardia	A09	Output Problem		
									A090811	Unable to Obtain Readings		
							E060110	Ventricular Fibrillation	A09	Output Problem		
									A090811	Unable to Obtain Readings		
							F02	Death	A09	Output Problem		
									A090811	Unable to Obtain Readings		
					CT2750	Implantable cardiac electrical stimulators and associated devices	E060109	Tachycardia	A09	Output Problem		
									A090811	Unable to Obtain Readings		
							E060110	Ventricular Fibrillation	A09	Output Problem		
									A090811	Unable to Obtain Readings		
							F02	Death	A09	Output Problem		
									A090811	Unable to Obtain Readings		
					CT2866	Implantable cardiac electrical stimulators	E060109	Tachycardia	A09	Output Problem		
									A090811	Unable to Obtain Readings		
							E060110	Ventricular Fibrillation	A09	Output Problem		
									A090811	Unable to Obtain Readings		
							F02	Death	A09	Output Problem		
									A090811	Unable to Obtain Readings		
					CT2985	Active	E060109	Tachycardia	A09	Output Problem		
									A090811	Unable to Obtain Readings		
							E060110	Ventricular Fibrillation	A09	Output Problem		
									A090811	Unable to Obtain Readings		
							F02	Death	A09	Output Problem		
									A090811	Unable to Obtain Readings		
					CT3001	Implantable metal	E060109	Tachycardia	A09	Output Problem		
									A090811	Unable to Obtain Readings		
							E060110	Ventricular Fibrillation	A09	Output Problem		
									A090811	Unable to Obtain Readings		
							F02	Death	A09	Output Problem		
									A090811	Unable to Obtain Readings		
					CT3002	Implantable synthetic polymer	E060109	Tachycardia	A09	Output Problem		
									A090811	Unable to Obtain Readings		
							E060110	Ventricular Fibrillation	A09	Output Problem		
									A090811	Unable to Obtain Readings		
							F02	Death	A09	Output Problem		
									A090811	Unable to Obtain Readings		
					CT3006	Implantable inorganic material	E060109	Tachycardia	A09	Output Problem		
									A090811	Unable to Obtain Readings		
							E060110	Ventricular Fibrillation	A09	Output Problem		
									A090811	Unable to Obtain Readings		
							F02	Death	A09	Output Problem		
									A090811	Unable to Obtain Readings		
					CT314	Electrically powered	E060109	Tachycardia	A09	Output Problem		
									A090811	Unable to Obtain Readings		
							E060110	Ventricular Fibrillation	A09	Output Problem		
									A090811	Unable to Obtain Readings		
							F02	Death	A09	Output Problem		
									A090811	Unable to Obtain Readings		
					CT321	Long-term surgical invasive	E060109	Tachycardia	A09	Output Problem		
									A090811	Unable to Obtain Readings		
							E060110	Ventricular Fibrillation	A09	Output Problem		
									A090811	Unable to Obtain Readings		
							F02	Death	A09	Output Problem		
									A090811	Unable to Obtain Readings		
					CT449	Generators and associated devices	E060109	Tachycardia	A09	Output Problem		
									A090811	Unable to Obtain Readings		

							E060110	Ventricular Fibrillation	A09	Output Problem		
									A090811	Unable to Obtain Readings		
							F02	Death	A09	Output Problem		
									A090811	Unable to Obtain Readings		
					CT981	Single use	E060109	Tachycardia	A09	Output Problem		
									A090811	Unable to Obtain Readings		
							E060110	Ventricular Fibrillation	A09	Output Problem		
									A090811	Unable to Obtain Readings		
							F02	Death	A09	Output Problem		
									A090811	Unable to Obtain Readings		
					CT983	Surgical invasive	E060109	Tachycardia	A09	Output Problem		
									A090811	Unable to Obtain Readings		
							E060110	Ventricular Fibrillation	A09	Output Problem		
									A090811	Unable to Obtain Readings		
							F02	Death	A09	Output Problem		
									A090811	Unable to Obtain Readings		
8071	2022	Death	17882	Physiologic-monitoring defibrillation system	CT1684	Cardiac pacemakers and associated devices	E0602	Cardiac Arrest	A071302	Inappropriate/Inadequate Shock/Stimulation		
							F02	Death	A071302	Inappropriate/Inadequate Shock/Stimulation		
8072	2019	All other reportable incidents	47265	Dual-chamber implantable pacemaker, rate-responsive	CT141	Electrodes/Leads and associated devices	F1905	Device Revision or Replacement	A0713	Defibrillation/Stimulation Problem	14/12/2018	
					CT145	Cardiovascular devices	F1905	Device Revision or Replacement	A0713	Defibrillation/Stimulation Problem	14/12/2018	
					CT1684	Cardiac pacemakers and associated devices	F1905	Device Revision or Replacement	A0713	Defibrillation/Stimulation Problem	14/12/2018	
					CT1685	Cardiac pacemakers	F1905	Device Revision or Replacement	A0713	Defibrillation/Stimulation Problem	14/12/2018	
					CT233	Surgical	F1905	Device Revision or Replacement	A0713	Defibrillation/Stimulation Problem	14/12/2018	
					CT2406	Implantable	F1905	Device Revision or Replacement	A0713	Defibrillation/Stimulation Problem	14/12/2018	
					CT2407	Active implantable	F1905	Device Revision or Replacement	A0713	Defibrillation/Stimulation Problem	14/12/2018	
					CT257	Implantable cardiac pacemakers	F1905	Device Revision or Replacement	A0713	Defibrillation/Stimulation Problem	14/12/2018	
					CT2727	Active implant devices/ancillaries	F1905	Device Revision or Replacement	A0713	Defibrillation/Stimulation Problem	14/12/2018	
					CT2750	Implantable cardiac electrical stimulators and associated devices	F1905	Device Revision or Replacement	A0713	Defibrillation/Stimulation Problem	14/12/2018	
					CT2866	Implantable cardiac electrical stimulators	F1905	Device Revision or Replacement	A0713	Defibrillation/Stimulation Problem	14/12/2018	
					CT2985	Active	F1905	Device Revision or Replacement	A0713	Defibrillation/Stimulation Problem	14/12/2018	
					CT3001	Implantable metal	F1905	Device Revision or Replacement	A0713	Defibrillation/Stimulation Problem	14/12/2018	
					CT3002	Implantable synthetic polymer	F1905	Device Revision or Replacement	A0713	Defibrillation/Stimulation Problem	14/12/2018	
					CT3006	Implantable inorganic material	F1905	Device Revision or Replacement	A0713	Defibrillation/Stimulation Problem	14/12/2018	
					CT314	Electrically powered	F1905	Device Revision or Replacement	A0713	Defibrillation/Stimulation Problem	14/12/2018	
					CT321	Long-term surgical invasive	F1905	Device Revision or Replacement	A0713	Defibrillation/Stimulation Problem	14/12/2018	
					CT449	Generators and associated devices	F1905	Device Revision or Replacement	A0713	Defibrillation/Stimulation Problem	14/12/2018	
					CT981	Single use	F1905	Device Revision or Replacement	A0713	Defibrillation/Stimulation Problem	14/12/2018	
					CT983	Surgical invasive	F1905	Device Revision or Replacement	A0713	Defibrillation/Stimulation Problem	14/12/2018	
8073	2019	All other reportable incidents	35852	Single-chamber implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	31/12/2018	
8074	2019	All other reportable incidents	35223	Endocardial/interventricular septal pacing lead	CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070909	Oversensing		
8075	2019	All other reportable incidents	37265	Dual-chamber implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	F1905	Device Revision or Replacement	A0705	Battery Problem	18/12/2018	
8076	2019	All other reportable incidents	35852	Single-chamber implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A070909	Oversensing		
8077	2019	All other reportable incidents	35852	Single-chamber implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A070909	Oversensing		
8078	2019	All other reportable incidents	35852	Single-chamber implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A0712	Pacing Problem		
8079	2019	All other reportable incidents	35852	Single-chamber implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A070909	Oversensing		
8080	2019	All other reportable incidents	36241	Pacemaker/defibrillator lead	CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A05	Mechanical Problem	22/10/2019	22/10/2019
8081	2019	All other reportable incidents	37265	Dual-chamber implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A070909	Oversensing		
8082	2019	All other reportable incidents	36241	Pacemaker/defibrillator lead	CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A072201	High Impedance		
8083	2019	All other reportable incidents	37265	Dual-chamber implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A072201	High Impedance		
8084	2019	All other reportable incidents	37265	Dual-chamber implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
8085	2019	All other reportable incidents	47265	Dual-chamber implantable pacemaker, rate-responsive	CT141	Electrodes/Leads and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem	25/11/2019	
					CT145	Cardiovascular devices	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem	25/11/2019	
					CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem	25/11/2019	
					CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem	25/11/2019	
					CT233	Surgical	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem	25/11/2019	
					CT2406	Implantable	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem	25/11/2019	
					CT2407	Active implantable	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem	25/11/2019	
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem	25/11/2019	
					CT2727	Active implant devices/ancillaries	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem	25/11/2019	
					CT2750	Implantable cardiac electrical stimulators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem	25/11/2019	
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem	25/11/2019	
					CT2985	Active	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem	25/11/2019	
					CT3001	Implantable metal	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem	25/11/2019	
					CT3002	Implantable synthetic polymer	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem	25/11/2019	
					CT3006	Implantable inorganic material	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem	25/11/2019	
					CT314	Electrically powered	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem	25/11/2019	
					CT321	Long-term surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem	25/11/2019	
					CT449	Generators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem	25/11/2019	
					CT981	Single use	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem	25/11/2019	
					CT983	Surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem	25/11/2019	
8086	2019	All other reportable incidents	35852	Single-chamber implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A071302	Inappropriate/Inadequate Shock/Stimulation		
8087	2019	All other reportable incidents	35852	Single-chamber implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A072201	High Impedance		
							F26	No Health Consequences or Impact	A072201	High Impedance		
8088	2019	All other reportable incidents	36241	Pacemaker/defibrillator lead	CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A07	Electrical /Electronic Property Problem		
8089	2019	All other reportable incidents	36241	Pacemaker/defibrillator lead	CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A072201	High Impedance		
8090	2019	All other reportable incidents	36241	Pacemaker/defibrillator lead	CT1684	Cardiac pacemakers and associated devices	F1905	Device Revision or Replacement	A	Annex A	23/10/2019	
8091	2019	All other reportable incidents	36241	Pacemaker/defibrillator lead	CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070909	Oversensing		
									A071302	Inappropriate/Inadequate Shock/Stimulation		
8092	2019	All other reportable incidents	35223	Endocardial/interventricular septal pacing lead	CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A072201	High Impedance		
8093	2019	All other reportable incidents	35852	Single-chamber implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A070909	Oversensing		
8094	2019	All other reportable incidents	37265	Dual-chamber implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A070909	Oversensing		
8095	2019	All other reportable incidents	37265	Dual-chamber implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	08/01/2019	
8096	2019	All other reportable incidents	35852	Single-chamber implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A070909	Oversensing		
									A071302	Inappropriate/Inadequate Shock/Stimulation		
8097	2019	All other reportable incidents	36241	Pacemaker/defibrillator lead	CT1684	Cardiac pacemakers and associated devices	F1905	Device Revision or Replacement	A072201	High Impedance		

8098	2019	All other reportable incidents	37265	Dual-chamber implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	F1905	Device Revision or Replacement	A072201	High Impedance	11/12/2018	
8099	2019	All other reportable incidents	47270	Cardiac resynchronization therapy implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A0722	Impedance Problem		
8100	2019	All other reportable incidents	35852	Single-chamber implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
8101	2019	All other reportable incidents	47205	Cardiac pulse generator programmer	CT1684	Cardiac pacemakers and associated devices	F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A110201	Application Program Freezes or Fails to Launch		
8102	2019	All other reportable incidents	36241	Pacemaker/defibrillator lead	CT1684	Cardiac pacemakers and associated devices	F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A010402	Migration		
							F1905	Device Revision or Replacement	A010402	Migration		
8103	2019	All other reportable incidents	37265	Dual-chamber implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	F1905	Device Revision or Replacement	A1205	Loose or Intermittent Connection	13/12/2018	
8104	2019	All other reportable incidents	36241	Pacemaker/defibrillator lead	CT1684	Cardiac pacemakers and associated devices	F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A070905	High Sensing Threshold	13/12/2018	
							F1905	Device Revision or Replacement	A070905	High Sensing Threshold	13/12/2018	
8105	2019	All other reportable incidents	47267	Single-chamber implantable pacemaker, rate-responsive	CT141	Electrodes/Leads and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem		
					CT145	Cardiovascular devices	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem		
					CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem		
					CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem		
					CT233	Surgical	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem		
					CT2406	Implantable	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem		
					CT2407	Active implantable	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem		
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem		
					CT2727	Active implant devices/ancillaries	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem		
					CT2750	Implantable cardiac electrical stimulators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem		
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem		
					CT2985	Active	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem		
					CT3001	Implantable metal	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem		
					CT3002	Implantable synthetic polymer	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem		
					CT3006	Implantable inorganic material	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem		
					CT314	Electrically powered	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem		
					CT321	Long-term surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem		
					CT449	Generators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem		
					CT981	Single use	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem		
					CT983	Surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem		
8106	2019	All other reportable incidents	37265	Dual-chamber implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	01/03/2019	
8107	2019	All other reportable incidents	47267	Single-chamber implantable pacemaker, rate-responsive	CT141	Electrodes/Leads and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
					CT145	Cardiovascular devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
					CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
					CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
					CT233	Surgical	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
					CT2406	Implantable	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
					CT2407	Active implantable	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
					CT2727	Active implant devices/ancillaries	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
					CT2750	Implantable cardiac electrical stimulators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
					CT2985	Active	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
					CT3001	Implantable metal	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
					CT3002	Implantable synthetic polymer	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
					CT3006	Implantable inorganic material	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
					CT314	Electrically powered	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
					CT321	Long-term surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
					CT449	Generators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
					CT981	Single use	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
					CT983	Surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
8108	2019	All other reportable incidents	35852	Single-chamber implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	25/10/2019	
8109	2019	All other reportable incidents	35223	Endocardial/interventricular septal pacing lead	CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A072202	Low Impedance		
8110	2019	All other reportable incidents	35223	Endocardial/interventricular septal pacing lead	CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A04	Material Integrity Problem		
8111	2019	All other reportable incidents	36241	Pacemaker/defibrillator lead	CT1684	Cardiac pacemakers and associated devices	F1905	Device Revision or Replacement	A072201	High Impedance	14/12/2018	
8112	2019	All other reportable incidents	35223	Endocardial/interventricular septal pacing lead	CT1684	Cardiac pacemakers and associated devices	F1905	Device Revision or Replacement	A070909	Oversensing		
8113	2019	All other reportable incidents	35852	Single-chamber implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	27/12/2018	
8114	2019	All other reportable incidents	37265	Dual-chamber implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
8115	2019	All other reportable incidents	37265	Dual-chamber implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
8116	2019	All other reportable incidents	35223	Endocardial/interventricular septal pacing lead	CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070909	Oversensing		
8117	2019	All other reportable incidents	35852	Single-chamber implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A070909	Oversensing		
8118	2019	All other reportable incidents	47270	Cardiac resynchronization therapy implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A	Annex A		
8119	2019	All other reportable incidents	37265	Dual-chamber implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
8120	2019	All other reportable incidents	37265	Dual-chamber implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A0709	Device Sensing Problem		
8121	2019	All other reportable incidents	47267	Single-chamber implantable pacemaker, rate-responsive	CT141	Electrodes/Leads and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem		
					CT145	Cardiovascular devices	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem		
					CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem		
					CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem		
					CT233	Surgical	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem		
					CT2406	Implantable	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem		
					CT2407	Active implantable	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem		
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem		
					CT2727	Active implant devices/ancillaries	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem		
					CT2750	Implantable cardiac electrical stimulators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem		
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem		
					CT2985	Active	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem		
					CT3001	Implantable metal	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem		
					CT3002	Implantable synthetic polymer	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem		
					CT3006	Implantable inorganic material	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem		
					CT314	Electrically powered	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem		
					CT321	Long-term surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem		
					CT449	Generators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem		

					CT981	Single use	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem		
					CT983	Surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem		
8122	2019	All other reportable incidents	36241	Pacemaker/defibrillator lead	CT1684	Cardiac pacemakers and associated devices	F1905	Device Revision or Replacement	A072201	High Impedance		
8123	2019	All other reportable incidents	36241	Pacemaker/defibrillator lead	CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A072201	High Impedance		
8124	2018	All other reportable incidents	35852	Single-chamber implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A071302	Inappropriate/Inadequate Shock/Stimulation		
8125	2019	All other reportable incidents	47267	Single-chamber implantable pacemaker, rate-responsive	CT141	Electrodes/Leads and associated devices	F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery		
					CT145	Cardiovascular devices	F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery		
					CT1684	Cardiac pacemakers and associated devices	F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery		
					CT1685	Cardiac pacemakers	F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery		
					CT233	Surgical	F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery		
					CT2406	Implantable	F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery		
					CT2407	Active implantable	F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery		
					CT257	Implantable cardiac pacemakers	F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery		
					CT2727	Active implant devices/ancillaries	F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery		
					CT2750	Implantable cardiac electrical stimulators and associated devices	F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery		
					CT2866	Implantable cardiac electrical stimulators	F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery		
					CT2985	Active	F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery		
					CT3001	Implantable metal	F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery		
					CT3002	Implantable synthetic polymer	F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery		
					CT3006	Implantable inorganic material	F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery		
					CT314	Electrically powered	F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery		
					CT321	Long-term surgical invasive	F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery		
					CT449	Generators and associated devices	F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery		
					CT981	Single use	F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery		
					CT983	Surgical invasive	F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery		
8126	2019	All other reportable incidents	37265	Dual-chamber implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A072201	High Impedance		
8127	2019	All other reportable incidents	35223	Endocardial/interventricular septal pacing lead	CT1684	Cardiac pacemakers and associated devices	F1905	Device Revision or Replacement	A040101	Fracture		
8128	2019	All other reportable incidents	35852	Single-chamber implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A071302	Inappropriate/Inadequate Shock/Stimulation		
8129	2019	All other reportable incidents	35852	Single-chamber implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A0709	Device Sensing Problem		
8130	2019	All other reportable incidents	37265	Dual-chamber implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	02/01/2019	
8131	2019	All other reportable incidents	37265	Dual-chamber implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	28/12/2018	
8132	2019	All other reportable incidents	35852	Single-chamber implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A070909	Oversensing		
8133	2019	All other reportable incidents	37265	Dual-chamber implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A072201	High Impedance		
8134	2019	All other reportable incidents	37265	Dual-chamber implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A070909	Oversensing		
8135	2019	All other reportable incidents	36241	Pacemaker/defibrillator lead	CT1684	Cardiac pacemakers and associated devices	F1905	Device Revision or Replacement	A070908	Failure to Sense	17/12/2018	
8136	2019	All other reportable incidents	47270	Cardiac resynchronization therapy implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	F1905	Device Revision or Replacement	A070101	Failure to Capture		
8137	2019	All other reportable incidents	47265	Dual-chamber implantable pacemaker, rate-responsive	CT141	Electrodes/Leads and associated devices	F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery		
					CT145	Cardiovascular devices	F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery		
					CT1684	Cardiac pacemakers and associated devices	F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery		
					CT1685	Cardiac pacemakers	F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery		
					CT233	Surgical	F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery		
					CT2406	Implantable	F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery		
					CT2407	Active implantable	F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery		
					CT257	Implantable cardiac pacemakers	F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery		
					CT2727	Active implant devices/ancillaries	F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery		
					CT2750	Implantable cardiac electrical stimulators and associated devices	F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery		
					CT2866	Implantable cardiac electrical stimulators	F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery		
					CT2985	Active	F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery		
					CT3001	Implantable metal	F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery		
					CT3002	Implantable synthetic polymer	F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery		
					CT3006	Implantable inorganic material	F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery		
					CT314	Electrically powered	F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery		
					CT321	Long-term surgical invasive	F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery		
					CT449	Generators and associated devices	F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery		
					CT981	Single use	F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery		
					CT983	Surgical invasive	F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery		
8138	2019	All other reportable incidents	36241	Pacemaker/defibrillator lead	CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A0722	Impedance Problem		
8139	2019	All other reportable incidents	36241	Pacemaker/defibrillator lead	CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A072201	High Impedance		
8140	2019	All other reportable incidents	47265	Dual-chamber implantable pacemaker, rate-responsive	CT141	Electrodes/Leads and associated devices	F1905	Device Revision or Replacement	A072201	High Impedance	14/12/2018	
					CT145	Cardiovascular devices	F1905	Device Revision or Replacement	A072201	High Impedance	14/12/2018	
					CT1684	Cardiac pacemakers and associated devices	F1905	Device Revision or Replacement	A072201	High Impedance	14/12/2018	
					CT1685	Cardiac pacemakers	F1905	Device Revision or Replacement	A072201	High Impedance	14/12/2018	
					CT233	Surgical	F1905	Device Revision or Replacement	A072201	High Impedance	14/12/2018	
					CT2406	Implantable	F1905	Device Revision or Replacement	A072201	High Impedance	14/12/2018	
					CT2407	Active implantable	F1905	Device Revision or Replacement	A072201	High Impedance	14/12/2018	
					CT257	Implantable cardiac pacemakers	F1905	Device Revision or Replacement	A072201	High Impedance	14/12/2018	
					CT2727	Active implant devices/ancillaries	F1905	Device Revision or Replacement	A072201	High Impedance	14/12/2018	
					CT2750	Implantable cardiac electrical stimulators and associated devices	F1905	Device Revision or Replacement	A072201	High Impedance	14/12/2018	
					CT2866	Implantable cardiac electrical stimulators	F1905	Device Revision or Replacement	A072201	High Impedance	14/12/2018	
					CT2985	Active	F1905	Device Revision or Replacement	A072201	High Impedance	14/12/2018	
					CT3001	Implantable metal	F1905	Device Revision or Replacement	A072201	High Impedance	14/12/2018	
					CT3002	Implantable synthetic polymer	F1905	Device Revision or Replacement	A072201	High Impedance	14/12/2018	
					CT3006	Implantable inorganic material	F1905	Device Revision or Replacement	A072201	High Impedance	14/12/2018	
					CT314	Electrically powered	F1905	Device Revision or Replacement	A072201	High Impedance	14/12/2018	
					CT321	Long-term surgical invasive	F1905	Device Revision or Replacement	A072201	High Impedance	14/12/2018	
					CT449	Generators and associated devices	F1905	Device Revision or Replacement	A072201	High Impedance	14/12/2018	
					CT981	Single use	F1905	Device Revision or Replacement	A072201	High Impedance	14/12/2018	
					CT983	Surgical invasive	F1905	Device Revision or Replacement	A072201	High Impedance	14/12/2018	
8141	2019	All other reportable incidents	37265	Dual-chamber implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A072201	High Impedance		
8142	2019	All other reportable incidents	37265	Dual-chamber implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	F1905	Device Revision or Replacement	A071102	Failure to Interrogate	08/01/2019	
8143	2019	All other reportable incidents	37265	Dual-chamber implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery		
8144	2019	All other reportable incidents	36241	Pacemaker/defibrillator lead	CT1684	Cardiac pacemakers and associated devices	F1905	Device Revision or Replacement	A010402	Migration		
8145	2019	All other reportable incidents	47265	Dual-chamber implantable pacemaker, rate-responsive	CT141	Electrodes/Leads and associated devices	F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery		
					CT145	Cardiovascular devices	F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery		

					CT1684	Cardiac pacemakers and associated devices	F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery		
					CT1685	Cardiac pacemakers	F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery		
					CT233	Surgical	F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery		
					CT2406	Implantable	F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery		
					CT2407	Active implantable	F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery		
					CT257	Implantable cardiac pacemakers	F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery		
					CT2727	Active implant devices/ancillaries	F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery		
					CT2750	Implantable cardiac electrical stimulators and associated devices	F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery		
					CT2866	Implantable cardiac electrical stimulators	F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery		
					CT2985	Active	F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery		
					CT3001	Implantable metal	F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery		
					CT3002	Implantable synthetic polymer	F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery		
					CT3006	Implantable inorganic material	F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery		
					CT314	Electrically powered	F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery		
					CT321	Long-term surgical invasive	F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery		
					CT449	Generators and associated devices	F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery		
					CT981	Single use	F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery		
					CT983	Surgical invasive	F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery		
8146	2019	All other reportable incidents	47265	Dual-chamber implantable pacemaker, rate-responsive	CT141	Electrodes/Leads and associated devices	F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	01/11/2019	
					CT145	Cardiovascular devices	F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	01/11/2019	
					CT1684	Cardiac pacemakers and associated devices	F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	01/11/2019	
					CT1685	Cardiac pacemakers	F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	01/11/2019	
					CT233	Surgical	F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	01/11/2019	
					CT2406	Implantable	F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	01/11/2019	
					CT2407	Active implantable	F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	01/11/2019	
					CT257	Implantable cardiac pacemakers	F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	01/11/2019	
					CT2727	Active implant devices/ancillaries	F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	01/11/2019	
					CT2750	Implantable cardiac electrical stimulators and associated devices	F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	01/11/2019	
					CT2866	Implantable cardiac electrical stimulators	F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	01/11/2019	
					CT2985	Active	F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	01/11/2019	
					CT3001	Implantable metal	F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	01/11/2019	
					CT3002	Implantable synthetic polymer	F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	01/11/2019	
					CT3006	Implantable inorganic material	F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	01/11/2019	
					CT314	Electrically powered	F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	01/11/2019	
					CT321	Long-term surgical invasive	F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	01/11/2019	
					CT449	Generators and associated devices	F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	01/11/2019	
					CT981	Single use	F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	01/11/2019	
					CT983	Surgical invasive	F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	01/11/2019	
8147	2019	All other reportable incidents	37265	Dual-chamber implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery		
8148	2019	All other reportable incidents	47265	Dual-chamber implantable pacemaker, rate-responsive	CT141	Electrodes/Leads and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
					CT145	Cardiovascular devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
					CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
					CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
					CT233	Surgical	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
					CT2406	Implantable	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
					CT2407	Active implantable	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
					CT2727	Active implant devices/ancillaries	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
					CT2750	Implantable cardiac electrical stimulators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
					CT2985	Active	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
					CT3001	Implantable metal	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
					CT3002	Implantable synthetic polymer	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
					CT3006	Implantable inorganic material	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
					CT314	Electrically powered	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
					CT321	Long-term surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
					CT449	Generators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
					CT981	Single use	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
					CT983	Surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
8149	2019	Death	35852	Single-chamber implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A070910	Undersensing		
8150	2019	All other reportable incidents	37265	Dual-chamber implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	07/01/2019	
8151	2019	All other reportable incidents	36241	Pacemaker/defibrillator lead	CT1684	Cardiac pacemakers and associated devices	F1905	Device Revision or Replacement	A010402	Migration	18/12/2018	
8152	2019	All other reportable incidents	47265	Dual-chamber implantable pacemaker, rate-responsive	CT141	Electrodes/Leads and associated devices	F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	13/07/2017	
					CT145	Cardiovascular devices	F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	13/07/2017	
					CT1684	Cardiac pacemakers and associated devices	F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	13/07/2017	
					CT1685	Cardiac pacemakers	F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	13/07/2017	
					CT233	Surgical	F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	13/07/2017	
					CT2406	Implantable	F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	13/07/2017	
					CT2407	Active implantable	F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	13/07/2017	
					CT257	Implantable cardiac pacemakers	F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	13/07/2017	
					CT2727	Active implant devices/ancillaries	F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	13/07/2017	
					CT2750	Implantable cardiac electrical stimulators and associated devices	F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	13/07/2017	
					CT2866	Implantable cardiac electrical stimulators	F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	13/07/2017	
					CT2985	Active	F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	13/07/2017	
					CT3001	Implantable metal	F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	13/07/2017	
					CT3002	Implantable synthetic polymer	F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	13/07/2017	
					CT3006	Implantable inorganic material	F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	13/07/2017	
					CT314	Electrically powered	F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	13/07/2017	
					CT321	Long-term surgical invasive	F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	13/07/2017	
					CT449	Generators and associated devices	F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	13/07/2017	
					CT981	Single use	F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	13/07/2017	
					CT983	Surgical invasive	F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	13/07/2017	
8153	2019	All other reportable incidents	37265	Dual-chamber implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A071302	Inappropriate/Inadequate Shock/Stimulation		
8154	2019	All other reportable incidents	36241	Pacemaker/defibrillator lead	CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A0510	Retraction Problem	28/12/2018	
8155	2019	All other reportable incidents	47265	Dual-chamber implantable pacemaker, rate-	CT141	Electrodes/Leads and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		

				responsive	CT145	Cardiovascular devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
					CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
					CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
					CT233	Surgical	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
					CT2406	Implantable	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
					CT2407	Active implantable	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
					CT2727	Active implant devices/ancillaries	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
					CT2750	Implantable cardiac electrical stimulators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
					CT2985	Active	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
					CT3001	Implantable metal	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
					CT3002	Implantable synthetic polymer	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
					CT3006	Implantable inorganic material	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
					CT314	Electrically powered	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
					CT321	Long-term surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
					CT449	Generators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
					CT981	Single use	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
					CT983	Surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
8156	2019	All other reportable incidents	37265	Dual-chamber implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
8157	2019	All other reportable incidents	37265	Dual-chamber implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	F1905	Device Revision or Replacement	A090201	Device Displays Incorrect Message	05/11/2019	
8158	2019	All other reportable incidents	35852	Single-chamber implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A070909	Oversensing		
8159	2019	All other reportable incidents	35852	Single-chamber implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A071302	Inappropriate/Inadequate Shock/Stimulation		
8160	2019	All other reportable incidents	35852	Single-chamber implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A0203	Defective Device		
8161	2019	All other reportable incidents	47265	Dual-chamber implantable pacemaker, rate-responsive	CT141	Electrodes/Leads and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
					CT145	Cardiovascular devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
					CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
					CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
					CT233	Surgical	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
					CT2406	Implantable	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
					CT2407	Active implantable	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
					CT2727	Active implant devices/ancillaries	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
					CT2750	Implantable cardiac electrical stimulators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
					CT2985	Active	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
					CT3001	Implantable metal	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
					CT3002	Implantable synthetic polymer	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
					CT3006	Implantable inorganic material	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
					CT314	Electrically powered	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
					CT321	Long-term surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
					CT449	Generators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
					CT981	Single use	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
					CT983	Surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
8162	2019	All other reportable incidents	47267	Single-chamber implantable pacemaker, rate-responsive	CT141	Electrodes/Leads and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
					CT145	Cardiovascular devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
					CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
					CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
					CT233	Surgical	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
					CT2406	Implantable	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
					CT2407	Active implantable	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
					CT2727	Active implant devices/ancillaries	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
					CT2750	Implantable cardiac electrical stimulators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
					CT2985	Active	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
					CT3001	Implantable metal	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
					CT3002	Implantable synthetic polymer	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
					CT3006	Implantable inorganic material	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
					CT314	Electrically powered	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
					CT321	Long-term surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
					CT449	Generators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
					CT981	Single use	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
					CT983	Surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
8163	2019	All other reportable incidents	35223	Endocardial/interventricular septal pacing lead	CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A072202	Low Impedance		
8164	2019	All other reportable incidents	47265	Dual-chamber implantable pacemaker, rate-responsive	CT141	Electrodes/Leads and associated devices	F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery		
					CT145	Cardiovascular devices	F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery		
					CT1684	Cardiac pacemakers and associated devices	F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery		
					CT1685	Cardiac pacemakers	F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery		
					CT233	Surgical	F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery		
					CT2406	Implantable	F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery		
					CT2407	Active implantable	F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery		
					CT257	Implantable cardiac pacemakers	F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery		
					CT2727	Active implant devices/ancillaries	F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery		
					CT2750	Implantable cardiac electrical stimulators and associated devices	F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery		
					CT2866	Implantable cardiac electrical stimulators	F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery		
					CT2985	Active	F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery		
					CT3001	Implantable metal	F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery		
					CT3002	Implantable synthetic polymer	F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery		
					CT3006	Implantable inorganic material	F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery		
					CT314	Electrically powered	F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery		
					CT321	Long-term surgical invasive	F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery		
					CT449	Generators and associated devices	F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery		
					CT981	Single use	F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery		

					CT983	Surgical invasive	F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery			
8165	2019	All other reportable incidents	47270	Cardiac resynchronization therapy implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	30/01/2019		
8166	2019	All other reportable incidents	47270	Cardiac resynchronization therapy implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	F1905	Device Revision or Replacement	A0705	Battery Problem	21/01/2019		
8167	2019	All other reportable incidents	36241	Pacemaker/defibrillator lead	CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A0510	Retraction Problem	09/01/2019		
8168	2019	All other reportable incidents	47265	Dual-chamber implantable pacemaker, rate-responsive	CT141	Electrodes/Leads and associated devices	F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery			
					CT145	Cardiovascular devices	F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery			
					CT1684	Cardiac pacemakers and associated devices	F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery			
					CT1685	Cardiac pacemakers	F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery			
					CT233	Surgical	F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery			
					CT2406	Implantable	F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery			
					CT2407	Active implantable	F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery			
					CT257	Implantable cardiac pacemakers	F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery			
					CT2727	Active implant devices/ancillaries	F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery			
					CT2750	Implantable cardiac electrical stimulators and associated devices	F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery			
					CT2866	Implantable cardiac electrical stimulators	F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery			
					CT2985	Active	F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery			
					CT3001	Implantable metal	F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery			
					CT3002	Implantable synthetic polymer	F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery			
					CT3006	Implantable inorganic material	F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery			
					CT314	Electrically powered	F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery			
					CT321	Long-term surgical invasive	F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery			
					CT449	Generators and associated devices	F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery			
					CT981	Single use	F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery			
					CT983	Surgical invasive	F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery			
8169	2019	All other reportable incidents	47263	Cardiac resynchronization therapy implantable pacemaker	CT1684	Cardiac pacemakers and associated devices	F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	16/01/2019		
					CT1685	Cardiac pacemakers	F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	16/01/2019		
					CT257	Implantable cardiac pacemakers	F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	16/01/2019		
					CT2866	Implantable cardiac electrical stimulators	F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	16/01/2019		
8170	2019	All other reportable incidents	47265	Dual-chamber implantable pacemaker, rate-responsive	CT141	Electrodes/Leads and associated devices	F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery			
					CT145	Cardiovascular devices	F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery			
					CT1684	Cardiac pacemakers and associated devices	F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery			
					CT1685	Cardiac pacemakers	F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery			
					CT233	Surgical	F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery			
					CT2406	Implantable	F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery			
					CT2407	Active implantable	F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery			
					CT257	Implantable cardiac pacemakers	F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery			
					CT2727	Active implant devices/ancillaries	F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery			
					CT2750	Implantable cardiac electrical stimulators and associated devices	F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery			
					CT2866	Implantable cardiac electrical stimulators	F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery			
					CT2985	Active	F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery			
					CT3001	Implantable metal	F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery			
					CT3002	Implantable synthetic polymer	F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery			
					CT3006	Implantable inorganic material	F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery			
					CT314	Electrically powered	F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery			
					CT321	Long-term surgical invasive	F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery			
					CT449	Generators and associated devices	F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery			
					CT981	Single use	F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery			
					CT983	Surgical invasive	F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery			
8171	2019	All other reportable incidents	37265	Dual-chamber implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	04/01/2019		
8172	2019	All other reportable incidents	37265	Dual-chamber implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	09/01/2019		
8173	2019	All other reportable incidents	47263	Cardiac resynchronization therapy implantable pacemaker	CT1684	Cardiac pacemakers and associated devices	F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery			
					CT1685	Cardiac pacemakers	F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery			
					CT257	Implantable cardiac pacemakers	F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery			
					CT2866	Implantable cardiac electrical stimulators	F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery			
8174	2019	All other reportable incidents	47265	Dual-chamber implantable pacemaker, rate-responsive	CT141	Electrodes/Leads and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A072201	High Impedance			
					CT145	Cardiovascular devices	E2403	No Clinical Signs, Symptoms or Conditions	A072201	High Impedance			
					CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A072201	High Impedance			
					CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A072201	High Impedance			
					CT233	Surgical	E2403	No Clinical Signs, Symptoms or Conditions	A072201	High Impedance			
					CT2406	Implantable	E2403	No Clinical Signs, Symptoms or Conditions	A072201	High Impedance			
					CT2407	Active implantable	E2403	No Clinical Signs, Symptoms or Conditions	A072201	High Impedance			
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A072201	High Impedance			
					CT2727	Active implant devices/ancillaries	E2403	No Clinical Signs, Symptoms or Conditions	A072201	High Impedance			
					CT2750	Implantable cardiac electrical stimulators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A072201	High Impedance			
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A072201	High Impedance			
					CT2985	Active	E2403	No Clinical Signs, Symptoms or Conditions	A072201	High Impedance			
					CT3001	Implantable metal	E2403	No Clinical Signs, Symptoms or Conditions	A072201	High Impedance			
					CT3002	Implantable synthetic polymer	E2403	No Clinical Signs, Symptoms or Conditions	A072201	High Impedance			
					CT3006	Implantable inorganic material	E2403	No Clinical Signs, Symptoms or Conditions	A072201	High Impedance			
					CT314	Electrically powered	E2403	No Clinical Signs, Symptoms or Conditions	A072201	High Impedance			
					CT321	Long-term surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A072201	High Impedance			
					CT449	Generators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A072201	High Impedance			
					CT981	Single use	E2403	No Clinical Signs, Symptoms or Conditions	A072201	High Impedance			
					CT983	Surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A072201	High Impedance			
8175	2019	All other reportable incidents	47270	Cardiac resynchronization therapy implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A090201	Device Displays Incorrect Message			
8176	2019	All other reportable incidents	37265	Dual-chamber implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A071302	Inappropriate/Inadequate Shock/Stimulation			
8177	2019	All other reportable incidents	47270	Cardiac resynchronization therapy implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	F1905	Device Revision or Replacement	A26	Insufficient Device Problem Information			
8178	2019	All other reportable incidents	37265	Dual-chamber implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	09/01/2019	09/01/2019	
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	09/01/2019	09/01/2019	
							F19	Surgical Intervention	A070504	Premature Discharge of Battery	09/01/2019	09/01/2019	
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	09/01/2019	09/01/2019	
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery	09/01/2019	09/01/2019	

8179	2019	All other reportable incidents	35852	Single-chamber implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A070909	Oversensing		
8180	2019	All other reportable incidents	47265	Dual-chamber implantable pacemaker, rate-responsive	CT141	Electrodes/Leads and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
					CT145	Cardiovascular devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
					CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
					CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
					CT233	Surgical	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
					CT2406	Implantable	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
					CT2407	Active implantable	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
					CT2727	Active implant devices/ancillaries	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
					CT2750	Implantable cardiac electrical stimulators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
					CT2985	Active	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
					CT3001	Implantable metal	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
					CT3002	Implantable synthetic polymer	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
					CT3006	Implantable inorganic material	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
					CT314	Electrically powered	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
					CT321	Long-term surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
					CT449	Generators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
					CT981	Single use	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
					CT983	Surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
8181	2019	All other reportable incidents	47265	Dual-chamber implantable pacemaker, rate-responsive	CT141	Electrodes/Leads and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A0711	Interrogation Problem		
					CT145	Cardiovascular devices	E2403	No Clinical Signs, Symptoms or Conditions	A0711	Interrogation Problem		
					CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A0711	Interrogation Problem		
					CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A0711	Interrogation Problem		
					CT233	Surgical	E2403	No Clinical Signs, Symptoms or Conditions	A0711	Interrogation Problem		
					CT2406	Implantable	E2403	No Clinical Signs, Symptoms or Conditions	A0711	Interrogation Problem		
					CT2407	Active implantable	E2403	No Clinical Signs, Symptoms or Conditions	A0711	Interrogation Problem		
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A0711	Interrogation Problem		
					CT2727	Active implant devices/ancillaries	E2403	No Clinical Signs, Symptoms or Conditions	A0711	Interrogation Problem		
					CT2750	Implantable cardiac electrical stimulators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A0711	Interrogation Problem		
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A0711	Interrogation Problem		
					CT2985	Active	E2403	No Clinical Signs, Symptoms or Conditions	A0711	Interrogation Problem		
					CT3001	Implantable metal	E2403	No Clinical Signs, Symptoms or Conditions	A0711	Interrogation Problem		
					CT3002	Implantable synthetic polymer	E2403	No Clinical Signs, Symptoms or Conditions	A0711	Interrogation Problem		
					CT3006	Implantable inorganic material	E2403	No Clinical Signs, Symptoms or Conditions	A0711	Interrogation Problem		
					CT314	Electrically powered	E2403	No Clinical Signs, Symptoms or Conditions	A0711	Interrogation Problem		
					CT321	Long-term surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A0711	Interrogation Problem		
					CT449	Generators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A0711	Interrogation Problem		
					CT981	Single use	E2403	No Clinical Signs, Symptoms or Conditions	A0711	Interrogation Problem		
					CT983	Surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A0711	Interrogation Problem		
8182	2019	All other reportable incidents	37265	Dual-chamber implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A070910	Undersensing		
8183	2019	All other reportable incidents	36241	Pacemaker/defibrillator lead	CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070101	Failure to Capture		
8184	2019	All other reportable incidents	37265	Dual-chamber implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	F1905	Device Revision or Replacement	A0705	Battery Problem	21/01/2019	
8185	2019	All other reportable incidents	37265	Dual-chamber implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A072201	High Impedance		
8186	2019	All other reportable incidents	35852	Single-chamber implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A070909	Oversensing		
8187	2019	All other reportable incidents	35852	Single-chamber implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A070909	Oversensing		
8188	2019	All other reportable incidents	36241	Pacemaker/defibrillator lead	CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A0510	Retraction Problem	30/11/2018	
8189	2019	All other reportable incidents	35852	Single-chamber implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A0709	Device Sensing Problem		
8190	2019	All other reportable incidents	36241	Pacemaker/defibrillator lead	CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A0709	Device Sensing Problem		
8191	2019	All other reportable incidents	36241	Pacemaker/defibrillator lead	CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A0709	Device Sensing Problem		
8192	2019	All other reportable incidents	37265	Dual-chamber implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A070909	Oversensing		
8193	2019	All other reportable incidents	47264	Dual-chamber implantable pacemaker, demand	CT141	Electrodes/Leads and associated devices	F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	01/03/2017	
					CT145	Cardiovascular devices	F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	01/03/2017	
					CT1684	Cardiac pacemakers and associated devices	F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	01/03/2017	
					CT1685	Cardiac pacemakers	F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	01/03/2017	
					CT233	Surgical	F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	01/03/2017	
					CT2406	Implantable	F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	01/03/2017	
					CT2407	Active implantable	F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	01/03/2017	
					CT257	Implantable cardiac pacemakers	F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	01/03/2017	
					CT2727	Active implant devices/ancillaries	F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	01/03/2017	
					CT2750	Implantable cardiac electrical stimulators and associated devices	F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	01/03/2017	
					CT2866	Implantable cardiac electrical stimulators	F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	01/03/2017	
					CT2985	Active	F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	01/03/2017	
					CT3001	Implantable metal	F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	01/03/2017	
					CT3002	Implantable synthetic polymer	F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	01/03/2017	
					CT3006	Implantable inorganic material	F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	01/03/2017	
					CT314	Electrically powered	F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	01/03/2017	
					CT321	Long-term surgical invasive	F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	01/03/2017	
					CT449	Generators and associated devices	F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	01/03/2017	
					CT981	Single use	F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	01/03/2017	
					CT983	Surgical invasive	F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	01/03/2017	
8194	2019	All other reportable incidents	35852	Single-chamber implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A070909	Oversensing		
8195	2019	All other reportable incidents	36241	Pacemaker/defibrillator lead	CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A072201	High Impedance		
8196	2019	All other reportable incidents	35852	Single-chamber implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
8197	2019	All other reportable incidents	37265	Dual-chamber implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery		
8198	2019	All other reportable incidents	47267	Single-chamber implantable pacemaker, rate-responsive	CT141	Electrodes/Leads and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
					CT145	Cardiovascular devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
					CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
					CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
					CT233	Surgical	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
					CT2406	Implantable	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
					CT2407	Active implantable	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
					CT2727	Active implant devices/ancillaries	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		

					CT2750	Implantable cardiac electrical stimulators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
					CT2985	Active	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
					CT3001	Implantable metal	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
					CT3002	Implantable synthetic polymer	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
					CT3006	Implantable inorganic material	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
					CT314	Electrically powered	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
					CT321	Long-term surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
					CT449	Generators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
					CT981	Single use	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
					CT983	Surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
8199	2019	All other reportable incidents	36241	Pacemaker/defibrillator lead	CT1684	Cardiac pacemakers and associated devices	E2401	Insufficient Information	A072201	High Impedance		
8200	2019	All other reportable incidents	35039	Epicardial pacing lead	CT1684	Cardiac pacemakers and associated devices	F1905	Device Revision or Replacement	A0405	Degraded		
8201	2019	All other reportable incidents	36241	Pacemaker/defibrillator lead	CT1684	Cardiac pacemakers and associated devices	F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A072201	High Impedance	24/01/2019	
8202	2019	Death	17882	Physiologic-monitoring defibrillation system	CT1684	Cardiac pacemakers and associated devices	F1001	Absence of Treatment	A1303	Intermittent Communication Failure		
8203	2019	All other reportable incidents	36241	Pacemaker/defibrillator lead	CT1684	Cardiac pacemakers and associated devices	F1905	Device Revision or Replacement	A0404	Crack		
8204	2019	All other reportable incidents	36241	Pacemaker/defibrillator lead	CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A0510	Retraction Problem		
8205	2019	All other reportable incidents	47267	Single-chamber implantable pacemaker, rate-responsive	CT141	Electrodes/Leads and associated devices	E2007	Fall	A0712	Pacing Problem		
							F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A0712	Pacing Problem		
					CT145	Cardiovascular devices	E2007	Fall	A0712	Pacing Problem		
							F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A0712	Pacing Problem		
					CT1684	Cardiac pacemakers and associated devices	E2007	Fall	A0712	Pacing Problem		
							F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A0712	Pacing Problem		
					CT1685	Cardiac pacemakers	E2007	Fall	A0712	Pacing Problem		
							F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A0712	Pacing Problem		
					CT233	Surgical	E2007	Fall	A0712	Pacing Problem		
							F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A0712	Pacing Problem		
					CT2406	Implantable	E2007	Fall	A0712	Pacing Problem		
							F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A0712	Pacing Problem		
					CT2407	Active implantable	E2007	Fall	A0712	Pacing Problem		
							F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A0712	Pacing Problem		
					CT257	Implantable cardiac pacemakers	E2007	Fall	A0712	Pacing Problem		
							F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A0712	Pacing Problem		
					CT2727	Active implant devices/ancillaries	E2007	Fall	A0712	Pacing Problem		
							F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A0712	Pacing Problem		
					CT2750	Implantable cardiac electrical stimulators and associated devices	E2007	Fall	A0712	Pacing Problem		
							F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A0712	Pacing Problem		
					CT2866	Implantable cardiac electrical stimulators	E2007	Fall	A0712	Pacing Problem		
							F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A0712	Pacing Problem		
					CT2985	Active	E2007	Fall	A0712	Pacing Problem		
							F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A0712	Pacing Problem		
					CT3001	Implantable metal	E2007	Fall	A0712	Pacing Problem		
							F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A0712	Pacing Problem		
					CT3002	Implantable synthetic polymer	E2007	Fall	A0712	Pacing Problem		
							F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A0712	Pacing Problem		
					CT3006	Implantable inorganic material	E2007	Fall	A0712	Pacing Problem		
							F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A0712	Pacing Problem		
					CT314	Electrically powered	E2007	Fall	A0712	Pacing Problem		
							F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A0712	Pacing Problem		
					CT321	Long-term surgical invasive	E2007	Fall	A0712	Pacing Problem		
							F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A0712	Pacing Problem		
					CT449	Generators and associated devices	E2007	Fall	A0712	Pacing Problem		
							F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A0712	Pacing Problem		
					CT981	Single use	E2007	Fall	A0712	Pacing Problem		
							F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A0712	Pacing Problem		
					CT983	Surgical invasive	E2007	Fall	A0712	Pacing Problem		
							F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A0712	Pacing Problem		
8206	2019	All other reportable incidents	35223	Endocardial/interventricular septal pacing lead	CT1684	Cardiac pacemakers and associated devices	F1905	Device Revision or Replacement	A070101	Failure to Capture		
8207	2019	All other reportable incidents	35223	Endocardial/interventricular septal pacing lead	CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A0414	Material Split, Cut or Torn	08/01/2019	
8208	2019	All other reportable incidents	35852	Single-chamber implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	F1905	Device Revision or Replacement	A070603	Failure to Charge	21/01/2019	
8209	2019	All other reportable incidents	37265	Dual-chamber implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A070909	Oversensing		
8210	2019	All other reportable incidents	47270	Cardiac resynchronization therapy implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E1906	Unspecified Infection	A010102	Device Appears to Trigger Rejection		
							F1905	Device Revision or Replacement	A010102	Device Appears to Trigger Rejection		
8211	2019	All other reportable incidents	36241	Pacemaker/defibrillator lead	CT1684	Cardiac pacemakers and associated devices	E1906	Unspecified Infection	A24	No Apparent Device Problem		
8212	2019	All other reportable incidents	37265	Dual-chamber implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A071302	Inappropriate/Inadequate Shock/Stimulation		
8213	2019	All other reportable incidents	36241	Pacemaker/defibrillator lead	CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A	Annex A		
8214	2019	All other reportable incidents	47940	Triple-chamber implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	24/01/2019	
8215	2019	All other reportable incidents	37265	Dual-chamber implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
8216	2019	All other reportable incidents	37265	Dual-chamber implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A070504	Premature Discharge of Battery		
8217	2019	All other reportable incidents	37265	Dual-chamber implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A070909	Oversensing		
8218	2019	All other reportable incidents	37265	Dual-chamber implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery		
8219	2019	All other reportable incidents	47267	Single-chamber implantable pacemaker, rate-responsive	CT141	Electrodes/Leads and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A09	Output Problem		
					CT145	Cardiovascular devices	E2403	No Clinical Signs, Symptoms or Conditions	A09	Output Problem		
					CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A09	Output Problem		
					CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A09	Output Problem		
					CT233	Surgical	E2403	No Clinical Signs, Symptoms or Conditions	A09	Output Problem		
					CT2406	Implantable	E2403	No Clinical Signs, Symptoms or Conditions	A09	Output Problem		
					CT2407	Active implantable	E2403	No Clinical Signs, Symptoms or Conditions	A09	Output Problem		
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A09	Output Problem		
					CT2727	Active implant devices/ancillaries	E2403	No Clinical Signs, Symptoms or Conditions	A09	Output Problem		
					CT2750	Implantable cardiac electrical stimulators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A09	Output Problem		
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A09	Output Problem		
					CT2985	Active	E2403	No Clinical Signs, Symptoms or Conditions	A09	Output Problem		
					CT3001	Implantable metal	E2403	No Clinical Signs, Symptoms or Conditions	A09	Output Problem		
					CT3002	Implantable synthetic polymer	E2403	No Clinical Signs, Symptoms or Conditions	A09	Output Problem		

					CT3006	Implantable inorganic material	E2403	No Clinical Signs, Symptoms or Conditions	A09	Output Problem		
					CT314	Electrically powered	E2403	No Clinical Signs, Symptoms or Conditions	A09	Output Problem		
					CT321	Long-term surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A09	Output Problem		
					CT449	Generators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A09	Output Problem		
					CT981	Single use	E2403	No Clinical Signs, Symptoms or Conditions	A09	Output Problem		
					CT983	Surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A09	Output Problem		
8220	2019	All other reportable incidents	47267	Single-chamber implantable pacemaker, rate-responsive	CT141	Electrodes/Leads and associated devices	F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	16/01/2019	
					CT145	Cardiovascular devices	F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	16/01/2019	
					CT1684	Cardiac pacemakers and associated devices	F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	16/01/2019	
					CT1685	Cardiac pacemakers	F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	16/01/2019	
					CT233	Surgical	F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	16/01/2019	
					CT2406	Implantable	F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	16/01/2019	
					CT2407	Active implantable	F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	16/01/2019	
					CT257	Implantable cardiac pacemakers	F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	16/01/2019	
					CT2727	Active implant devices/ancillaries	F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	16/01/2019	
					CT2750	Implantable cardiac electrical stimulators and associated devices	F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	16/01/2019	
					CT2866	Implantable cardiac electrical stimulators	F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	16/01/2019	
					CT2985	Active	F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	16/01/2019	
					CT3001	Implantable metal	F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	16/01/2019	
					CT3002	Implantable synthetic polymer	F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	16/01/2019	
					CT3006	Implantable inorganic material	F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	16/01/2019	
					CT314	Electrically powered	F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	16/01/2019	
					CT321	Long-term surgical invasive	F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	16/01/2019	
					CT449	Generators and associated devices	F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	16/01/2019	
					CT981	Single use	F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	16/01/2019	
					CT983	Surgical invasive	F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	16/01/2019	
8221	2019	All other reportable incidents	37265	Dual-chamber implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
8222	2019	All other reportable incidents	47263	Cardiac resynchronization therapy implantable pacemaker	CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A1305	Wireless Communication Problem		
					CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A1305	Wireless Communication Problem		
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A1305	Wireless Communication Problem		
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A1305	Wireless Communication Problem		
8223	2019	All other reportable incidents	37265	Dual-chamber implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A070909	Oversensing		
8224	2019	All other reportable incidents	35852	Single-chamber implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
8225	2019	All other reportable incidents	47270	Cardiac resynchronization therapy implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E2104	Electric Shock	A071302	Inappropriate/Inadequate Shock/Stimulation		
8226	2019	All other reportable incidents	35852	Single-chamber implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
8227	2019	All other reportable incidents	17882	Physiologic-monitoring defibrillation system	CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070803	Failure to Power Up		
8228	2019	Unanticipated serious deterioration in state of health	37265	Dual-chamber implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E1906	Unspecified Infection	A010102	Device Appears to Trigger Rejection	23/01/2019	
8229	2019	All other reportable incidents	35852	Single-chamber implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	F1905	Device Revision or Replacement	A070909	Oversensing	02/01/2019	
8230	2019	All other reportable incidents	16995	Intracardiac electrode	CT1684	Cardiac pacemakers and associated devices	F21	Unexpected Deterioration	A24	No Apparent Device Problem		
8231	2019	All other reportable incidents	35223	Endocardial/interventricular septal pacing lead	CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070102	High Capture Threshold		
8232	2019	All other reportable incidents	47265	Dual-chamber implantable pacemaker, rate-responsive	CT141	Electrodes/Leads and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
					CT145	Cardiovascular devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
					CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
					CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
					CT233	Surgical	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
					CT2406	Implantable	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
					CT2407	Active implantable	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
					CT2727	Active implant devices/ancillaries	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
					CT2750	Implantable cardiac electrical stimulators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
					CT2985	Active	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
					CT3001	Implantable metal	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
					CT3002	Implantable synthetic polymer	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
					CT3006	Implantable inorganic material	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
					CT314	Electrically powered	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
					CT321	Long-term surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
					CT449	Generators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
					CT981	Single use	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
					CT983	Surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
8233	2019	All other reportable incidents	47267	Single-chamber implantable pacemaker, rate-responsive	CT141	Electrodes/Leads and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
					CT145	Cardiovascular devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
					CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
					CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
					CT233	Surgical	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
					CT2406	Implantable	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
					CT2407	Active implantable	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
					CT2727	Active implant devices/ancillaries	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
					CT2750	Implantable cardiac electrical stimulators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
					CT2985	Active	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
					CT3001	Implantable metal	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
					CT3002	Implantable synthetic polymer	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
					CT3006	Implantable inorganic material	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
					CT314	Electrically powered	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
					CT321	Long-term surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
					CT449	Generators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
					CT981	Single use	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
					CT983	Surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
8234	2019	All other reportable incidents	47265	Dual-chamber implantable pacemaker, rate-responsive	CT141	Electrodes/Leads and associated devices	F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery		
					CT145	Cardiovascular devices	F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery		
					CT1684	Cardiac pacemakers and associated devices	F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery		

					CT1685	Cardiac pacemakers	F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery		
					CT233	Surgical	F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery		
					CT2406	Implantable	F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery		
					CT2407	Active implantable	F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery		
					CT257	Implantable cardiac pacemakers	F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery		
					CT2727	Active implant devices/ancillaries	F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery		
					CT2750	Implantable cardiac electrical stimulators and associated devices	F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery		
					CT2866	Implantable cardiac electrical stimulators	F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery		
					CT2985	Active	F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery		
					CT3001	Implantable metal	F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery		
					CT3002	Implantable synthetic polymer	F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery		
					CT3006	Implantable inorganic material	F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery		
					CT314	Electrically powered	F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery		
					CT321	Long-term surgical invasive	F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery		
					CT449	Generators and associated devices	F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery		
					CT981	Single use	F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery		
					CT983	Surgical invasive	F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery		
8235	2019	All other reportable incidents	47267	Single-chamber implantable pacemaker, rate-responsive	CT141	Electrodes/Leads and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070909	Oversensing		
					CT145	Cardiovascular devices	E2403	No Clinical Signs, Symptoms or Conditions	A070909	Oversensing		
					CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070909	Oversensing		
					CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A070909	Oversensing		
					CT233	Surgical	E2403	No Clinical Signs, Symptoms or Conditions	A070909	Oversensing		
					CT2406	Implantable	E2403	No Clinical Signs, Symptoms or Conditions	A070909	Oversensing		
					CT2407	Active implantable	E2403	No Clinical Signs, Symptoms or Conditions	A070909	Oversensing		
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A070909	Oversensing		
					CT2727	Active implant devices/ancillaries	E2403	No Clinical Signs, Symptoms or Conditions	A070909	Oversensing		
					CT2750	Implantable cardiac electrical stimulators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070909	Oversensing		
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A070909	Oversensing		
					CT2985	Active	E2403	No Clinical Signs, Symptoms or Conditions	A070909	Oversensing		
					CT3001	Implantable metal	E2403	No Clinical Signs, Symptoms or Conditions	A070909	Oversensing		
					CT3002	Implantable synthetic polymer	E2403	No Clinical Signs, Symptoms or Conditions	A070909	Oversensing		
					CT3006	Implantable inorganic material	E2403	No Clinical Signs, Symptoms or Conditions	A070909	Oversensing		
					CT314	Electrically powered	E2403	No Clinical Signs, Symptoms or Conditions	A070909	Oversensing		
					CT321	Long-term surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A070909	Oversensing		
					CT449	Generators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070909	Oversensing		
					CT981	Single use	E2403	No Clinical Signs, Symptoms or Conditions	A070909	Oversensing		
					CT983	Surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A070909	Oversensing		
8236	2019	All other reportable incidents	35852	Single-chamber implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	F1905	Device Revision or Replacement	A070603	Failure to Charge		
8237	2019	Unanticipated serious deterioration in state of health	36241	Pacemaker/defibrillator lead	CT1684	Cardiac pacemakers and associated devices	E1906	Unspecified Infection	A0101	Patient-Device Incompatibility	23/01/2019	
							F1905	Device Revision or Replacement	A0101	Patient-Device Incompatibility	23/01/2019	
8238	2019	All other reportable incidents	37265	Dual-chamber implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A072201	High Impedance		
8239	2019	All other reportable incidents	37265	Dual-chamber implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery		
8240	2019	All other reportable incidents	47270	Cardiac resynchronization therapy implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A070909	Oversensing		
8241	2019	All other reportable incidents	36241	Pacemaker/defibrillator lead	CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070908	Failure to Sense		
8242	2019	All other reportable incidents	47270	Cardiac resynchronization therapy implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A070102	High Capture Threshold		
8243	2019	All other reportable incidents	36241	Pacemaker/defibrillator lead	CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A0510	Retraction Problem	14/01/2019	
8244	2019	All other reportable incidents	47265	Dual-chamber implantable pacemaker, rate-responsive	CT141	Electrodes/Leads and associated devices	F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	31/01/2019	
					CT145	Cardiovascular devices	F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	31/01/2019	
					CT1684	Cardiac pacemakers and associated devices	F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	31/01/2019	
					CT1685	Cardiac pacemakers	F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	31/01/2019	
					CT233	Surgical	F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	31/01/2019	
					CT2406	Implantable	F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	31/01/2019	
					CT2407	Active implantable	F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	31/01/2019	
					CT257	Implantable cardiac pacemakers	F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	31/01/2019	
					CT2727	Active implant devices/ancillaries	F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	31/01/2019	
					CT2750	Implantable cardiac electrical stimulators and associated devices	F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	31/01/2019	
					CT2866	Implantable cardiac electrical stimulators	F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	31/01/2019	
					CT2985	Active	F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	31/01/2019	
					CT3001	Implantable metal	F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	31/01/2019	
					CT3002	Implantable synthetic polymer	F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	31/01/2019	
					CT3006	Implantable inorganic material	F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	31/01/2019	
					CT314	Electrically powered	F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	31/01/2019	
					CT321	Long-term surgical invasive	F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	31/01/2019	
					CT449	Generators and associated devices	F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	31/01/2019	
					CT981	Single use	F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	31/01/2019	
					CT983	Surgical invasive	F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	31/01/2019	
8245	2019	All other reportable incidents	35852	Single-chamber implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	29/01/2019	
8246	2019	All other reportable incidents	35852	Single-chamber implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	F1905	Device Revision or Replacement	A071102	Failure to Interrogate	17/01/2019	
8247	2019	All other reportable incidents	47265	Dual-chamber implantable pacemaker, rate-responsive	CT141	Electrodes/Leads and associated devices	F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	22/01/2019	
					CT145	Cardiovascular devices	F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	22/01/2019	
					CT1684	Cardiac pacemakers and associated devices	F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	22/01/2019	
					CT1685	Cardiac pacemakers	F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	22/01/2019	
					CT233	Surgical	F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	22/01/2019	
					CT2406	Implantable	F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	22/01/2019	
					CT2407	Active implantable	F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	22/01/2019	
					CT257	Implantable cardiac pacemakers	F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	22/01/2019	
					CT2727	Active implant devices/ancillaries	F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	22/01/2019	
					CT2750	Implantable cardiac electrical stimulators and associated devices	F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	22/01/2019	
					CT2866	Implantable cardiac electrical stimulators	F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	22/01/2019	
					CT2985	Active	F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	22/01/2019	
					CT3001	Implantable metal	F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	22/01/2019	
					CT3002	Implantable synthetic polymer	F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	22/01/2019	
					CT3006	Implantable inorganic material	F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	22/01/2019	

					CT314	Electrically powered	F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	22/01/2019	
					CT321	Long-term surgical invasive	F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	22/01/2019	
					CT449	Generators and associated devices	F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	22/01/2019	
					CT981	Single use	F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	22/01/2019	
					CT983	Surgical invasive	F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	22/01/2019	
8248	2019	All other reportable incidents	36241	Pacemaker/defibrillator lead	CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A0510	Retraction Problem	22/01/2019	
8249	2019	All other reportable incidents	37265	Dual-chamber implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	23/01/2019	
8250	2019	All other reportable incidents	36241	Pacemaker/defibrillator lead	CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A0510	Retraction Problem	14/01/2019	
8251	2019	All other reportable incidents	47267	Single-chamber implantable pacemaker, rate-responsive	CT141	Electrodes/Leads and associated devices	F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	20/11/2019	
					CT145	Cardiovascular devices	F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	20/11/2019	
					CT1684	Cardiac pacemakers and associated devices	F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	20/11/2019	
					CT1685	Cardiac pacemakers	F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	20/11/2019	
					CT233	Surgical	F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	20/11/2019	
					CT2406	Implantable	F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	20/11/2019	
					CT2407	Active implantable	F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	20/11/2019	
					CT257	Implantable cardiac pacemakers	F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	20/11/2019	
					CT2727	Active implant devices/ancillaries	F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	20/11/2019	
					CT2750	Implantable cardiac electrical stimulators and associated devices	F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	20/11/2019	
					CT2866	Implantable cardiac electrical stimulators	F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	20/11/2019	
					CT2985	Active	F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	20/11/2019	
					CT3001	Implantable metal	F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	20/11/2019	
					CT3002	Implantable synthetic polymer	F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	20/11/2019	
					CT3006	Implantable inorganic material	F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	20/11/2019	
					CT314	Electrically powered	F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	20/11/2019	
					CT321	Long-term surgical invasive	F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	20/11/2019	
					CT449	Generators and associated devices	F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	20/11/2019	
					CT981	Single use	F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	20/11/2019	
					CT983	Surgical invasive	F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	20/11/2019	
8252	2019	All other reportable incidents	47265	Dual-chamber implantable pacemaker, rate-responsive	CT141	Electrodes/Leads and associated devices	F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A072201	High Impedance		
					CT145	Cardiovascular devices	F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A072201	High Impedance		
					CT1684	Cardiac pacemakers and associated devices	F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A072201	High Impedance		
					CT1685	Cardiac pacemakers	F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A072201	High Impedance		
					CT233	Surgical	F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A072201	High Impedance		
					CT2406	Implantable	F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A072201	High Impedance		
					CT2407	Active implantable	F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A072201	High Impedance		
					CT257	Implantable cardiac pacemakers	F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A072201	High Impedance		
					CT2727	Active implant devices/ancillaries	F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A072201	High Impedance		
					CT2750	Implantable cardiac electrical stimulators and associated devices	F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A072201	High Impedance		
					CT2866	Implantable cardiac electrical stimulators	F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A072201	High Impedance		
					CT2985	Active	F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A072201	High Impedance		
					CT3001	Implantable metal	F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A072201	High Impedance		
					CT3002	Implantable synthetic polymer	F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A072201	High Impedance		
					CT3006	Implantable inorganic material	F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A072201	High Impedance		
					CT314	Electrically powered	F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A072201	High Impedance		
					CT321	Long-term surgical invasive	F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A072201	High Impedance		
					CT449	Generators and associated devices	F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A072201	High Impedance		
					CT981	Single use	F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A072201	High Impedance		
					CT983	Surgical invasive	F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A072201	High Impedance		
8253	2019	All other reportable incidents	47267	Single-chamber implantable pacemaker, rate-responsive	CT141	Electrodes/Leads and associated devices	F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery		
					CT145	Cardiovascular devices	F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery		
					CT1684	Cardiac pacemakers and associated devices	F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery		
					CT1685	Cardiac pacemakers	F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery		
					CT233	Surgical	F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery		
					CT2406	Implantable	F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery		
					CT2407	Active implantable	F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery		
					CT257	Implantable cardiac pacemakers	F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery		
					CT2727	Active implant devices/ancillaries	F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery		
					CT2750	Implantable cardiac electrical stimulators and associated devices	F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery		
					CT2866	Implantable cardiac electrical stimulators	F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery		
					CT2985	Active	F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery		
					CT3001	Implantable metal	F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery		
					CT3002	Implantable synthetic polymer	F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery		
					CT3006	Implantable inorganic material	F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery		
					CT314	Electrically powered	F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery		
					CT321	Long-term surgical invasive	F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery		
					CT449	Generators and associated devices	F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery		
					CT981	Single use	F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery		
					CT983	Surgical invasive	F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery		
8254	2019	All other reportable incidents	37265	Dual-chamber implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery		
8255	2019	All other reportable incidents	35852	Single-chamber implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A070909	Oversensing		
8256	2019	All other reportable incidents	47265	Dual-chamber implantable pacemaker, rate-responsive	CT141	Electrodes/Leads and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
					CT145	Cardiovascular devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
					CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
					CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
					CT233	Surgical	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
					CT2406	Implantable	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
					CT2407	Active implantable	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
					CT2727	Active implant devices/ancillaries	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
					CT2750	Implantable cardiac electrical stimulators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
					CT2985	Active	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
					CT3001	Implantable metal	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
					CT3002	Implantable synthetic polymer	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
					CT3006	Implantable inorganic material	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		

					CT314	Electrically powered	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
					CT321	Long-term surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
					CT449	Generators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
					CT981	Single use	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
					CT983	Surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
8257	2019	All other reportable incidents	47267	Single-chamber implantable pacemaker, rate-responsive	CT141	Electrodes/Leads and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
					CT145	Cardiovascular devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
					CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
					CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
					CT233	Surgical	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
					CT2406	Implantable	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
					CT2407	Active implantable	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
					CT2727	Active implant devices/ancillaries	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
					CT2750	Implantable cardiac electrical stimulators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
					CT2985	Active	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
					CT3001	Implantable metal	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
					CT3002	Implantable synthetic polymer	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
					CT3006	Implantable inorganic material	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
					CT314	Electrically powered	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
					CT321	Long-term surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
					CT449	Generators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
					CT981	Single use	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
					CT983	Surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
8258	2019	All other reportable incidents	47267	Single-chamber implantable pacemaker, rate-responsive	CT141	Electrodes/Leads and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
					CT145	Cardiovascular devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
					CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
					CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
					CT233	Surgical	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
					CT2406	Implantable	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
					CT2407	Active implantable	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
					CT2727	Active implant devices/ancillaries	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
					CT2750	Implantable cardiac electrical stimulators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
					CT2985	Active	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
					CT3001	Implantable metal	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
					CT3002	Implantable synthetic polymer	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
					CT3006	Implantable inorganic material	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
					CT314	Electrically powered	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
					CT321	Long-term surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
					CT449	Generators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
					CT981	Single use	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
					CT983	Surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
8259	2019	All other reportable incidents	12913	Pacemaker, implantable	CT145	Cardiovascular devices	E0506	Hemorrhage/Blood Loss/Bleeding	A1502	Positioning Problem		
							E0515	Vascular Dissection	A1502	Positioning Problem		
							F1901	Additional Surgery	A1502	Positioning Problem		
							F1903	Device Explantation	A1502	Positioning Problem		
							F21	Unexpected Deterioration	A1502	Positioning Problem		
					CT1684	Cardiac pacemakers and associated devices	E0506	Hemorrhage/Blood Loss/Bleeding	A1502	Positioning Problem		
							E0515	Vascular Dissection	A1502	Positioning Problem		
							F1901	Additional Surgery	A1502	Positioning Problem		
							F1903	Device Explantation	A1502	Positioning Problem		
							F21	Unexpected Deterioration	A1502	Positioning Problem		
					CT1685	Cardiac pacemakers	E0506	Hemorrhage/Blood Loss/Bleeding	A1502	Positioning Problem		
							E0515	Vascular Dissection	A1502	Positioning Problem		
							F1901	Additional Surgery	A1502	Positioning Problem		
							F1903	Device Explantation	A1502	Positioning Problem		
							F21	Unexpected Deterioration	A1502	Positioning Problem		
					CT2406	Implantable	E0506	Hemorrhage/Blood Loss/Bleeding	A1502	Positioning Problem		
							E0515	Vascular Dissection	A1502	Positioning Problem		
							F1901	Additional Surgery	A1502	Positioning Problem		
							F1903	Device Explantation	A1502	Positioning Problem		
							F21	Unexpected Deterioration	A1502	Positioning Problem		
					CT2407	Active implantable	E0506	Hemorrhage/Blood Loss/Bleeding	A1502	Positioning Problem		
							E0515	Vascular Dissection	A1502	Positioning Problem		
							F1901	Additional Surgery	A1502	Positioning Problem		
							F1903	Device Explantation	A1502	Positioning Problem		
							F21	Unexpected Deterioration	A1502	Positioning Problem		
					CT257	Implantable cardiac pacemakers	E0506	Hemorrhage/Blood Loss/Bleeding	A1502	Positioning Problem		
							E0515	Vascular Dissection	A1502	Positioning Problem		
							F1901	Additional Surgery	A1502	Positioning Problem		
							F1903	Device Explantation	A1502	Positioning Problem		
							F21	Unexpected Deterioration	A1502	Positioning Problem		
					CT2727	Active implant devices/ancillaries	E0506	Hemorrhage/Blood Loss/Bleeding	A1502	Positioning Problem		
							E0515	Vascular Dissection	A1502	Positioning Problem		
							F1901	Additional Surgery	A1502	Positioning Problem		
							F1903	Device Explantation	A1502	Positioning Problem		
							F21	Unexpected Deterioration	A1502	Positioning Problem		
					CT2750	Implantable cardiac electrical stimulators and associated devices	E0506	Hemorrhage/Blood Loss/Bleeding	A1502	Positioning Problem		
							E0515	Vascular Dissection	A1502	Positioning Problem		
							F1901	Additional Surgery	A1502	Positioning Problem		
							F1903	Device Explantation	A1502	Positioning Problem		
							F21	Unexpected Deterioration	A1502	Positioning Problem		
					CT2866	Implantable cardiac electrical stimulators	E0506	Hemorrhage/Blood Loss/Bleeding	A1502	Positioning Problem		

							E0515	Vascular Dissection	A1502	Positioning Problem		
							F1901	Additional Surgery	A1502	Positioning Problem		
							F1903	Device Explantation	A1502	Positioning Problem		
							F21	Unexpected Deterioration	A1502	Positioning Problem		
					CT2985	Active	E0506	Hemorrhage/Blood Loss/Bleeding	A1502	Positioning Problem		
							E0515	Vascular Dissection	A1502	Positioning Problem		
							F1901	Additional Surgery	A1502	Positioning Problem		
							F1903	Device Explantation	A1502	Positioning Problem		
							F21	Unexpected Deterioration	A1502	Positioning Problem		
					CT314	Electrically powered	E0506	Hemorrhage/Blood Loss/Bleeding	A1502	Positioning Problem		
							E0515	Vascular Dissection	A1502	Positioning Problem		
							F1901	Additional Surgery	A1502	Positioning Problem		
							F1903	Device Explantation	A1502	Positioning Problem		
							F21	Unexpected Deterioration	A1502	Positioning Problem		
8260	2019	All other reportable incidents	47270	Cardiac resynchronization therapy implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	31/01/2019	
8261	2019	All other reportable incidents	47265	Dual-chamber implantable pacemaker, rate-responsive	CT141	Electrodes/Leads and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070909	Oversensing		
					CT145	Cardiovascular devices	E2403	No Clinical Signs, Symptoms or Conditions	A070909	Oversensing		
					CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070909	Oversensing		
					CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A070909	Oversensing		
					CT233	Surgical	E2403	No Clinical Signs, Symptoms or Conditions	A070909	Oversensing		
					CT2406	Implantable	E2403	No Clinical Signs, Symptoms or Conditions	A070909	Oversensing		
					CT2407	Active implantable	E2403	No Clinical Signs, Symptoms or Conditions	A070909	Oversensing		
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A070909	Oversensing		
					CT2727	Active implant devices/ancillaries	E2403	No Clinical Signs, Symptoms or Conditions	A070909	Oversensing		
					CT2750	Implantable cardiac electrical stimulators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070909	Oversensing		
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A070909	Oversensing		
					CT2985	Active	E2403	No Clinical Signs, Symptoms or Conditions	A070909	Oversensing		
					CT3001	Implantable metal	E2403	No Clinical Signs, Symptoms or Conditions	A070909	Oversensing		
					CT3002	Implantable synthetic polymer	E2403	No Clinical Signs, Symptoms or Conditions	A070909	Oversensing		
					CT3006	Implantable inorganic material	E2403	No Clinical Signs, Symptoms or Conditions	A070909	Oversensing		
					CT314	Electrically powered	E2403	No Clinical Signs, Symptoms or Conditions	A070909	Oversensing		
					CT321	Long-term surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A070909	Oversensing		
					CT449	Generators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070909	Oversensing		
					CT981	Single use	E2403	No Clinical Signs, Symptoms or Conditions	A070909	Oversensing		
					CT983	Surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A070909	Oversensing		
8262	2019	All other reportable incidents	35852	Single-chamber implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A1304	Telemetry Discrepancy		
8263	2019	All other reportable incidents	36241	Pacemaker/defibrillator lead	CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A0510	Retraction Problem	23/01/2019	
8264	2019	All other reportable incidents	37265	Dual-chamber implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	01/02/2019	
8265	2019	All other reportable incidents	36241	Pacemaker/defibrillator lead	CT1684	Cardiac pacemakers and associated devices	F1905	Device Revision or Replacement	A072201	High Impedance		
8266	2019	All other reportable incidents	36241	Pacemaker/defibrillator lead	CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A0709	Device Sensing Problem		
8267	2019	All other reportable incidents	37265	Dual-chamber implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A070909	Oversensing		
8268	2019	All other reportable incidents	37265	Dual-chamber implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A071302	Inappropriate/Inadequate Shock/Stimulation		
8269	2019	All other reportable incidents	36241	Pacemaker/defibrillator lead	CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A0413	Material Separation		
8270	2019	All other reportable incidents	37265	Dual-chamber implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A071302	Inappropriate/Inadequate Shock/Stimulation		
8271	2019	All other reportable incidents	35852	Single-chamber implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A072201	High Impedance		
8272	2019	All other reportable incidents	47270	Cardiac resynchronization therapy implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
8273	2019	All other reportable incidents	35223	Endocardial/interventricular septal pacing lead	CT1684	Cardiac pacemakers and associated devices	E2114	Perforation	A0104	Migration or Expulsion of Device	30/01/2019	
							F1905	Device Revision or Replacement	A0104	Migration or Expulsion of Device	30/01/2019	
8274	2019	All other reportable incidents	36241	Pacemaker/defibrillator lead	CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070909	Oversensing		
8275	2019	All other reportable incidents	35852	Single-chamber implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
8276	2019	All other reportable incidents	47265	Dual-chamber implantable pacemaker, rate-responsive	CT141	Electrodes/Leads and associated devices	F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A070910	Undersensing		
					CT145	Cardiovascular devices	F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A070910	Undersensing		
					CT1684	Cardiac pacemakers and associated devices	F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A070910	Undersensing		
					CT1685	Cardiac pacemakers	F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A070910	Undersensing		
					CT233	Surgical	F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A070910	Undersensing		
					CT2406	Implantable	F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A070910	Undersensing		
					CT2407	Active implantable	F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A070910	Undersensing		
					CT257	Implantable cardiac pacemakers	F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A070910	Undersensing		
					CT2727	Active implant devices/ancillaries	F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A070910	Undersensing		
					CT2750	Implantable cardiac electrical stimulators and associated devices	F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A070910	Undersensing		
					CT2866	Implantable cardiac electrical stimulators	F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A070910	Undersensing		
					CT2985	Active	F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A070910	Undersensing		
					CT3001	Implantable metal	F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A070910	Undersensing		
					CT3002	Implantable synthetic polymer	F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A070910	Undersensing		
					CT3006	Implantable inorganic material	F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A070910	Undersensing		
					CT314	Electrically powered	F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A070910	Undersensing		
					CT321	Long-term surgical invasive	F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A070910	Undersensing		
					CT449	Generators and associated devices	F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A070910	Undersensing		
					CT981	Single use	F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A070910	Undersensing		
					CT983	Surgical invasive	F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A070910	Undersensing		
8277	2019	All other reportable incidents	35852	Single-chamber implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A071302	Inappropriate/Inadequate Shock/Stimulation		
8278	2019	All other reportable incidents	35852	Single-chamber implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
8279	2019	All other reportable incidents	47265	Dual-chamber implantable pacemaker, rate-responsive	CT141	Electrodes/Leads and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
					CT145	Cardiovascular devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
					CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
					CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
					CT233	Surgical	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
					CT2406	Implantable	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
					CT2407	Active implantable	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
					CT2727	Active implant devices/ancillaries	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
					CT2750	Implantable cardiac electrical stimulators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
					CT2985	Active	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		

					CT3001	Implantable metal	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
					CT3002	Implantable synthetic polymer	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
					CT3006	Implantable inorganic material	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
					CT314	Electrically powered	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
					CT321	Long-term surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
					CT449	Generators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
					CT981	Single use	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
					CT983	Surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
8280	2019	All other reportable incidents	36241	Pacemaker/defibrillator lead	CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070908	Failure to Sense		
8281	2019	All other reportable incidents	35852	Single-chamber implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A070909	Oversensing		
8282	2019	All other reportable incidents	36241	Pacemaker/defibrillator lead	CT1684	Cardiac pacemakers and associated devices	F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A071302	Inappropriate/Inadequate Shock/Stimulation		
8283	2019	All other reportable incidents	37265	Dual-chamber implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery		
8284	2019	All other reportable incidents	47263	Cardiac resynchronization therapy implantable pacemaker	CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
					CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
8285	2019	All other reportable incidents	35852	Single-chamber implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A070909	Oversensing		
8286	2019	All other reportable incidents	35852	Single-chamber implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A070909	Oversensing		
							F1905	Device Revision or Replacement	A070909	Oversensing		
8287	2019	All other reportable incidents	35223	Endocardial/interventricular septal pacing lead	CT1684	Cardiac pacemakers and associated devices	F1905	Device Revision or Replacement	A070909	Oversensing	03/12/2019	
							F21	Unexpected Deterioration	A070909	Oversensing	03/12/2019	
8288	2019	All other reportable incidents	36241	Pacemaker/defibrillator lead	CT1684	Cardiac pacemakers and associated devices	F1905	Device Revision or Replacement	A0510	Retraction Problem	23/01/2019	
8289	2019	All other reportable incidents	35223	Endocardial/interventricular septal pacing lead	CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A0709	Device Sensing Problem		
8290	2019	All other reportable incidents	47270	Cardiac resynchronization therapy implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A071302	Inappropriate/Inadequate Shock/Stimulation		
8291	2019	All other reportable incidents	35852	Single-chamber implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A070909	Oversensing		
8292	2019	All other reportable incidents	47265	Dual-chamber implantable pacemaker, rate-responsive	CT141	Electrodes/Leads and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070909	Oversensing		
					CT145	Cardiovascular devices	E2403	No Clinical Signs, Symptoms or Conditions	A070909	Oversensing		
					CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070909	Oversensing		
					CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A070909	Oversensing		
					CT233	Surgical	E2403	No Clinical Signs, Symptoms or Conditions	A070909	Oversensing		
					CT2406	Implantable	E2403	No Clinical Signs, Symptoms or Conditions	A070909	Oversensing		
					CT2407	Active implantable	E2403	No Clinical Signs, Symptoms or Conditions	A070909	Oversensing		
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A070909	Oversensing		
					CT2727	Active implant devices/ancillaries	E2403	No Clinical Signs, Symptoms or Conditions	A070909	Oversensing		
					CT2750	Implantable cardiac electrical stimulators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070909	Oversensing		
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A070909	Oversensing		
					CT2985	Active	E2403	No Clinical Signs, Symptoms or Conditions	A070909	Oversensing		
					CT3001	Implantable metal	E2403	No Clinical Signs, Symptoms or Conditions	A070909	Oversensing		
					CT3002	Implantable synthetic polymer	E2403	No Clinical Signs, Symptoms or Conditions	A070909	Oversensing		
					CT3006	Implantable inorganic material	E2403	No Clinical Signs, Symptoms or Conditions	A070909	Oversensing		
					CT314	Electrically powered	E2403	No Clinical Signs, Symptoms or Conditions	A070909	Oversensing		
					CT321	Long-term surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A070909	Oversensing		
					CT449	Generators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070909	Oversensing		
					CT981	Single use	E2403	No Clinical Signs, Symptoms or Conditions	A070909	Oversensing		
					CT983	Surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A070909	Oversensing		
8293	2019	All other reportable incidents	35852	Single-chamber implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	F21	Unexpected Deterioration	A090402	Failure to Deliver Energy		
8294	2019	All other reportable incidents	17882	Physiologic-monitoring defibrillation system	CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A0720	Electromagnetic Compatibility Problem		
8295	2019	All other reportable incidents	47265	Dual-chamber implantable pacemaker, rate-responsive	CT141	Electrodes/Leads and associated devices	F1905	Device Revision or Replacement	A0705	Battery Problem		
					CT145	Cardiovascular devices	F1905	Device Revision or Replacement	A0705	Battery Problem		
					CT1684	Cardiac pacemakers and associated devices	F1905	Device Revision or Replacement	A0705	Battery Problem		
					CT1685	Cardiac pacemakers	F1905	Device Revision or Replacement	A0705	Battery Problem		
					CT233	Surgical	F1905	Device Revision or Replacement	A0705	Battery Problem		
					CT2406	Implantable	F1905	Device Revision or Replacement	A0705	Battery Problem		
					CT2407	Active implantable	F1905	Device Revision or Replacement	A0705	Battery Problem		
					CT257	Implantable cardiac pacemakers	F1905	Device Revision or Replacement	A0705	Battery Problem		
					CT2727	Active implant devices/ancillaries	F1905	Device Revision or Replacement	A0705	Battery Problem		
					CT2750	Implantable cardiac electrical stimulators and associated devices	F1905	Device Revision or Replacement	A0705	Battery Problem		
					CT2866	Implantable cardiac electrical stimulators	F1905	Device Revision or Replacement	A0705	Battery Problem		
					CT2985	Active	F1905	Device Revision or Replacement	A0705	Battery Problem		
					CT3001	Implantable metal	F1905	Device Revision or Replacement	A0705	Battery Problem		
					CT3002	Implantable synthetic polymer	F1905	Device Revision or Replacement	A0705	Battery Problem		
					CT3006	Implantable inorganic material	F1905	Device Revision or Replacement	A0705	Battery Problem		
					CT314	Electrically powered	F1905	Device Revision or Replacement	A0705	Battery Problem		
					CT321	Long-term surgical invasive	F1905	Device Revision or Replacement	A0705	Battery Problem		
					CT449	Generators and associated devices	F1905	Device Revision or Replacement	A0705	Battery Problem		
					CT981	Single use	F1905	Device Revision or Replacement	A0705	Battery Problem		
					CT983	Surgical invasive	F1905	Device Revision or Replacement	A0705	Battery Problem		
8296	2019	All other reportable incidents	47270	Cardiac resynchronization therapy implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	F1905	Device Revision or Replacement	A0705	Battery Problem	06/02/2019	
8297	2019	All other reportable incidents	47265	Dual-chamber implantable pacemaker, rate-responsive	CT141	Electrodes/Leads and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070909	Oversensing		
					CT145	Cardiovascular devices	E2403	No Clinical Signs, Symptoms or Conditions	A070909	Oversensing		
					CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070909	Oversensing		
					CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A070909	Oversensing		
					CT233	Surgical	E2403	No Clinical Signs, Symptoms or Conditions	A070909	Oversensing		
					CT2406	Implantable	E2403	No Clinical Signs, Symptoms or Conditions	A070909	Oversensing		
					CT2407	Active implantable	E2403	No Clinical Signs, Symptoms or Conditions	A070909	Oversensing		
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A070909	Oversensing		
					CT2727	Active implant devices/ancillaries	E2403	No Clinical Signs, Symptoms or Conditions	A070909	Oversensing		
					CT2750	Implantable cardiac electrical stimulators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070909	Oversensing		
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A070909	Oversensing		
					CT2985	Active	E2403	No Clinical Signs, Symptoms or Conditions	A070909	Oversensing		
					CT3001	Implantable metal	E2403	No Clinical Signs, Symptoms or Conditions	A070909	Oversensing		
					CT3002	Implantable synthetic polymer	E2403	No Clinical Signs, Symptoms or Conditions	A070909	Oversensing		
					CT3006	Implantable inorganic material	E2403	No Clinical Signs, Symptoms or Conditions	A070909	Oversensing		

					CT314	Electrically powered	E2403	No Clinical Signs, Symptoms or Conditions	A070909	Oversensing		
					CT321	Long-term surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A070909	Oversensing		
					CT449	Generators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070909	Oversensing		
					CT981	Single use	E2403	No Clinical Signs, Symptoms or Conditions	A070909	Oversensing		
					CT983	Surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A070909	Oversensing		
8298	2019	All other reportable incidents	36241	Pacemaker/defibrillator lead	CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070909	Oversensing		
8299	2019	All other reportable incidents	37265	Dual-chamber implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	F21	Unexpected Deterioration	A070910	Undersensing		
8300	2019	Death	35223	Endocardial/interventricular septal pacing lead	CT1684	Cardiac pacemakers and associated devices	F21	Unexpected Deterioration	A0712	Pacing Problem		
8301	2019	All other reportable incidents	36241	Pacemaker/defibrillator lead	CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070909	Oversensing		
8302	2019	All other reportable incidents	47265	Dual-chamber implantable pacemaker, rate-responsive	CT141	Electrodes/Leads and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
					CT145	Cardiovascular devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
					CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
					CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
					CT233	Surgical	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
					CT2406	Implantable	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
					CT2407	Active implantable	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
					CT2727	Active implant devices/ancillaries	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
					CT2750	Implantable cardiac electrical stimulators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
					CT2985	Active	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
					CT3001	Implantable metal	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
					CT3002	Implantable synthetic polymer	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
					CT3006	Implantable inorganic material	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
					CT314	Electrically powered	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
					CT321	Long-term surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
					CT449	Generators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
					CT981	Single use	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
					CT983	Surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
8303	2019	All other reportable incidents	47263	Cardiac resynchronization therapy implantable pacemaker	CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
					CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
8304	2019	All other reportable incidents	47265	Dual-chamber implantable pacemaker, rate-responsive	CT141	Electrodes/Leads and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
					CT145	Cardiovascular devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
					CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
					CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
					CT233	Surgical	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
					CT2406	Implantable	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
					CT2407	Active implantable	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
					CT2727	Active implant devices/ancillaries	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
					CT2750	Implantable cardiac electrical stimulators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
					CT2985	Active	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
					CT3001	Implantable metal	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
					CT3002	Implantable synthetic polymer	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
					CT3006	Implantable inorganic material	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
					CT314	Electrically powered	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
					CT321	Long-term surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
					CT449	Generators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
					CT981	Single use	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
					CT983	Surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
8305	2019	All other reportable incidents	47263	Cardiac resynchronization therapy implantable pacemaker	CT1684	Cardiac pacemakers and associated devices	F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	06/02/2019	
					CT1685	Cardiac pacemakers	F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	06/02/2019	
					CT257	Implantable cardiac pacemakers	F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	06/02/2019	
					CT2866	Implantable cardiac electrical stimulators	F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	06/02/2019	
8306	2019	All other reportable incidents	36241	Pacemaker/defibrillator lead	CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A0510	Retraction Problem	28/01/2019	
8307	2019	All other reportable incidents	47265	Dual-chamber implantable pacemaker, rate-responsive	CT141	Electrodes/Leads and associated devices	F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	31/01/2019	
					CT145	Cardiovascular devices	F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	31/01/2019	
					CT1684	Cardiac pacemakers and associated devices	F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	31/01/2019	
					CT1685	Cardiac pacemakers	F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	31/01/2019	
					CT233	Surgical	F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	31/01/2019	
					CT2406	Implantable	F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	31/01/2019	
					CT2407	Active implantable	F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	31/01/2019	
					CT257	Implantable cardiac pacemakers	F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	31/01/2019	
					CT2727	Active implant devices/ancillaries	F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	31/01/2019	
					CT2750	Implantable cardiac electrical stimulators and associated devices	F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	31/01/2019	
					CT2866	Implantable cardiac electrical stimulators	F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	31/01/2019	
					CT2985	Active	F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	31/01/2019	
					CT3001	Implantable metal	F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	31/01/2019	
					CT3002	Implantable synthetic polymer	F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	31/01/2019	
					CT3006	Implantable inorganic material	F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	31/01/2019	
					CT314	Electrically powered	F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	31/01/2019	
					CT321	Long-term surgical invasive	F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	31/01/2019	
					CT449	Generators and associated devices	F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	31/01/2019	
					CT981	Single use	F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	31/01/2019	
					CT983	Surgical invasive	F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	31/01/2019	
8308	2019	All other reportable incidents	47265	Dual-chamber implantable pacemaker, rate-responsive	CT141	Electrodes/Leads and associated devices	F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	31/01/2019	
					CT145	Cardiovascular devices	F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	31/01/2019	
					CT1684	Cardiac pacemakers and associated devices	F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	31/01/2019	
					CT1685	Cardiac pacemakers	F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	31/01/2019	
					CT233	Surgical	F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	31/01/2019	
					CT2406	Implantable	F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	31/01/2019	
					CT2407	Active implantable	F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	31/01/2019	

					CT257	Implantable cardiac pacemakers	F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	31/01/2019		
					CT2727	Active implant devices/ancillaries	F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	31/01/2019		
					CT2750	Implantable cardiac electrical stimulators and associated devices	F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	31/01/2019		
					CT2866	Implantable cardiac electrical stimulators	F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	31/01/2019		
					CT2985	Active	F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	31/01/2019		
					CT3001	Implantable metal	F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	31/01/2019		
					CT3002	Implantable synthetic polymer	F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	31/01/2019		
					CT3006	Implantable inorganic material	F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	31/01/2019		
					CT314	Electrically powered	F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	31/01/2019		
					CT321	Long-term surgical invasive	F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	31/01/2019		
					CT449	Generators and associated devices	F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	31/01/2019		
					CT981	Single use	F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	31/01/2019		
					CT983	Surgical invasive	F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	31/01/2019		
8309	2019	All other reportable incidents	37265	Dual-chamber implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem			
8310	2019	All other reportable incidents	35852	Single-chamber implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A1304	Telemetry Discrepancy			
8311	2019	All other reportable incidents	36241	Pacemaker/defibrillator lead	CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A0510	Retraction Problem			
8312	2019	All other reportable incidents	37265	Dual-chamber implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	24/01/2019		
8313	2019	All other reportable incidents	35852	Single-chamber implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A070909	Oversensing			
8314	2019	All other reportable incidents	47267	Single-chamber implantable pacemaker, rate-responsive	CT141	Electrodes/Leads and associated devices	F1905	Device Revision or Replacement	A010102	Device Appears to Trigger Rejection	28/11/2018		
					CT145	Cardiovascular devices	F1905	Device Revision or Replacement	A010102	Device Appears to Trigger Rejection	28/11/2018		
					CT1684	Cardiac pacemakers and associated devices	F1905	Device Revision or Replacement	A010102	Device Appears to Trigger Rejection	28/11/2018		
					CT1685	Cardiac pacemakers	F1905	Device Revision or Replacement	A010102	Device Appears to Trigger Rejection	28/11/2018		
					CT233	Surgical	F1905	Device Revision or Replacement	A010102	Device Appears to Trigger Rejection	28/11/2018		
					CT2406	Implantable	F1905	Device Revision or Replacement	A010102	Device Appears to Trigger Rejection	28/11/2018		
					CT2407	Active implantable	F1905	Device Revision or Replacement	A010102	Device Appears to Trigger Rejection	28/11/2018		
					CT257	Implantable cardiac pacemakers	F1905	Device Revision or Replacement	A010102	Device Appears to Trigger Rejection	28/11/2018		
					CT2727	Active implant devices/ancillaries	F1905	Device Revision or Replacement	A010102	Device Appears to Trigger Rejection	28/11/2018		
					CT2750	Implantable cardiac electrical stimulators and associated devices	F1905	Device Revision or Replacement	A010102	Device Appears to Trigger Rejection	28/11/2018		
					CT2866	Implantable cardiac electrical stimulators	F1905	Device Revision or Replacement	A010102	Device Appears to Trigger Rejection	28/11/2018		
					CT2985	Active	F1905	Device Revision or Replacement	A010102	Device Appears to Trigger Rejection	28/11/2018		
					CT3001	Implantable metal	F1905	Device Revision or Replacement	A010102	Device Appears to Trigger Rejection	28/11/2018		
					CT3002	Implantable synthetic polymer	F1905	Device Revision or Replacement	A010102	Device Appears to Trigger Rejection	28/11/2018		
					CT3006	Implantable inorganic material	F1905	Device Revision or Replacement	A010102	Device Appears to Trigger Rejection	28/11/2018		
					CT314	Electrically powered	F1905	Device Revision or Replacement	A010102	Device Appears to Trigger Rejection	28/11/2018		
					CT321	Long-term surgical invasive	F1905	Device Revision or Replacement	A010102	Device Appears to Trigger Rejection	28/11/2018		
					CT449	Generators and associated devices	F1905	Device Revision or Replacement	A010102	Device Appears to Trigger Rejection	28/11/2018		
					CT981	Single use	F1905	Device Revision or Replacement	A010102	Device Appears to Trigger Rejection	28/11/2018		
					CT983	Surgical invasive	F1905	Device Revision or Replacement	A010102	Device Appears to Trigger Rejection	28/11/2018		
8315	2019	All other reportable incidents	36241	Pacemaker/defibrillator lead	CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A0510	Retraction Problem	18/01/2019		
8316	2019	All other reportable incidents	47267	Single-chamber implantable pacemaker, rate-responsive	CT141	Electrodes/Leads and associated devices	F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	28/01/2019		
					CT145	Cardiovascular devices	F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	28/01/2019		
					CT1684	Cardiac pacemakers and associated devices	F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	28/01/2019		
					CT1685	Cardiac pacemakers	F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	28/01/2019		
					CT233	Surgical	F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	28/01/2019		
					CT2406	Implantable	F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	28/01/2019		
					CT2407	Active implantable	F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	28/01/2019		
					CT257	Implantable cardiac pacemakers	F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	28/01/2019		
					CT2727	Active implant devices/ancillaries	F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	28/01/2019		
					CT2750	Implantable cardiac electrical stimulators and associated devices	F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	28/01/2019		
					CT2866	Implantable cardiac electrical stimulators	F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	28/01/2019		
					CT2985	Active	F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	28/01/2019		
					CT3001	Implantable metal	F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	28/01/2019		
					CT3002	Implantable synthetic polymer	F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	28/01/2019		
					CT3006	Implantable inorganic material	F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	28/01/2019		
					CT314	Electrically powered	F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	28/01/2019		
					CT321	Long-term surgical invasive	F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	28/01/2019		
					CT449	Generators and associated devices	F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	28/01/2019		
					CT981	Single use	F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	28/01/2019		
					CT983	Surgical invasive	F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	28/01/2019		
8317	2019	All other reportable incidents	37265	Dual-chamber implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A070909	Oversensing			
8318	2019	All other reportable incidents	36241	Pacemaker/defibrillator lead	CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A0510	Retraction Problem	13/02/2019		
8319	2019	All other reportable incidents	35852	Single-chamber implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A0709	Device Sensing Problem			
8320	2019	All other reportable incidents	36241	Pacemaker/defibrillator lead	CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A0510	Retraction Problem	13/02/2019		
8321	2019	All other reportable incidents	35852	Single-chamber implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	06/02/2019		
8322	2019	All other reportable incidents	37265	Dual-chamber implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E2104	Electric Shock	A071302	Inappropriate/Inadequate Shock/Stimulation			
							F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A071302	Inappropriate/Inadequate Shock/Stimulation			
8323	2019	All other reportable incidents	37265	Dual-chamber implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A070101	Failure to Capture			
8324	2019	Death	47265	Dual-chamber implantable pacemaker, rate-responsive	CT141	Electrodes/Leads and associated devices	E2401	Insufficient Information	A26	Insufficient Device Problem Information	06/02/2019		
					CT145	Cardiovascular devices	E2401	Insufficient Information	A26	Insufficient Device Problem Information	06/02/2019		
					CT1684	Cardiac pacemakers and associated devices	E2401	Insufficient Information	A26	Insufficient Device Problem Information	06/02/2019		
					CT1685	Cardiac pacemakers	E2401	Insufficient Information	A26	Insufficient Device Problem Information	06/02/2019		
					CT233	Surgical	E2401	Insufficient Information	A26	Insufficient Device Problem Information	06/02/2019		
					CT2406	Implantable	E2401	Insufficient Information	A26	Insufficient Device Problem Information	06/02/2019		
					CT2407	Active implantable	E2401	Insufficient Information	A26	Insufficient Device Problem Information	06/02/2019		
					CT257	Implantable cardiac pacemakers	E2401	Insufficient Information	A26	Insufficient Device Problem Information	06/02/2019		
					CT2727	Active implant devices/ancillaries	E2401	Insufficient Information	A26	Insufficient Device Problem Information	06/02/2019		
					CT2750	Implantable cardiac electrical stimulators and associated devices	E2401	Insufficient Information	A26	Insufficient Device Problem Information	06/02/2019		
					CT2866	Implantable cardiac electrical stimulators	E2401	Insufficient Information	A26	Insufficient Device Problem Information	06/02/2019		
					CT2985	Active	E2401	Insufficient Information	A26	Insufficient Device Problem Information	06/02/2019		
					CT3001	Implantable metal	E2401	Insufficient Information	A26	Insufficient Device Problem Information	06/02/2019		
					CT3002	Implantable synthetic polymer	E2401	Insufficient Information	A26	Insufficient Device Problem Information	06/02/2019		
					CT3006	Implantable inorganic material	E2401	Insufficient Information	A26	Insufficient Device Problem Information	06/02/2019		
					CT314	Electrically powered	E2401	Insufficient Information	A26	Insufficient Device Problem Information	06/02/2019		
					CT321	Long-term surgical invasive	E2401	Insufficient Information	A26	Insufficient Device Problem Information	06/02/2019		
					CT449	Generators and associated devices	E2401	Insufficient Information	A26	Insufficient Device Problem Information	06/02/2019		

					CT981	Single use	E2401	Insufficient Information	A26	Insufficient Device Problem Information	06/02/2019		
					CT983	Surgical invasive	E2401	Insufficient Information	A26	Insufficient Device Problem Information	06/02/2019		
8325	2020	All other reportable incidents	47265	Dual-chamber implantable pacemaker, rate-responsive	CT141	Electrodes/Leads and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	23/12/2019	23/12/2019	
									A071102	Failure to Interrogate	23/12/2019	23/12/2019	
							F1901	Additional Surgery	A070504	Premature Discharge of Battery	23/12/2019	23/12/2019	
									A071102	Failure to Interrogate	23/12/2019	23/12/2019	
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	23/12/2019	23/12/2019	
									A071102	Failure to Interrogate	23/12/2019	23/12/2019	
					CT145	Cardiovascular devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	23/12/2019	23/12/2019	
									A071102	Failure to Interrogate	23/12/2019	23/12/2019	
							F1901	Additional Surgery	A070504	Premature Discharge of Battery	23/12/2019	23/12/2019	
									A071102	Failure to Interrogate	23/12/2019	23/12/2019	
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	23/12/2019	23/12/2019	
									A071102	Failure to Interrogate	23/12/2019	23/12/2019	
					CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	23/12/2019	23/12/2019	
									A071102	Failure to Interrogate	23/12/2019	23/12/2019	
							F1901	Additional Surgery	A070504	Premature Discharge of Battery	23/12/2019	23/12/2019	
									A071102	Failure to Interrogate	23/12/2019	23/12/2019	
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	23/12/2019	23/12/2019	
									A071102	Failure to Interrogate	23/12/2019	23/12/2019	
					CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	23/12/2019	23/12/2019	
									A071102	Failure to Interrogate	23/12/2019	23/12/2019	
							F1901	Additional Surgery	A070504	Premature Discharge of Battery	23/12/2019	23/12/2019	
									A071102	Failure to Interrogate	23/12/2019	23/12/2019	
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	23/12/2019	23/12/2019	
									A071102	Failure to Interrogate	23/12/2019	23/12/2019	
					CT233	Surgical	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	23/12/2019	23/12/2019	
									A071102	Failure to Interrogate	23/12/2019	23/12/2019	
							F1901	Additional Surgery	A070504	Premature Discharge of Battery	23/12/2019	23/12/2019	
									A071102	Failure to Interrogate	23/12/2019	23/12/2019	
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	23/12/2019	23/12/2019	
									A071102	Failure to Interrogate	23/12/2019	23/12/2019	
					CT2406	Implantable	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	23/12/2019	23/12/2019	
									A071102	Failure to Interrogate	23/12/2019	23/12/2019	
							F1901	Additional Surgery	A070504	Premature Discharge of Battery	23/12/2019	23/12/2019	
									A071102	Failure to Interrogate	23/12/2019	23/12/2019	
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	23/12/2019	23/12/2019	
									A071102	Failure to Interrogate	23/12/2019	23/12/2019	
					CT2407	Active implantable	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	23/12/2019	23/12/2019	
									A071102	Failure to Interrogate	23/12/2019	23/12/2019	
							F1901	Additional Surgery	A070504	Premature Discharge of Battery	23/12/2019	23/12/2019	
									A071102	Failure to Interrogate	23/12/2019	23/12/2019	
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	23/12/2019	23/12/2019	
									A071102	Failure to Interrogate	23/12/2019	23/12/2019	
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	23/12/2019	23/12/2019	
									A071102	Failure to Interrogate	23/12/2019	23/12/2019	
							F1901	Additional Surgery	A070504	Premature Discharge of Battery	23/12/2019	23/12/2019	
									A071102	Failure to Interrogate	23/12/2019	23/12/2019	
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	23/12/2019	23/12/2019	
									A071102	Failure to Interrogate	23/12/2019	23/12/2019	
					CT2727	Active implant devices/ancillaries	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	23/12/2019	23/12/2019	
									A071102	Failure to Interrogate	23/12/2019	23/12/2019	
							F1901	Additional Surgery	A070504	Premature Discharge of Battery	23/12/2019	23/12/2019	
									A071102	Failure to Interrogate	23/12/2019	23/12/2019	
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	23/12/2019	23/12/2019	
									A071102	Failure to Interrogate	23/12/2019	23/12/2019	
					CT2750	Implantable cardiac electrical stimulators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	23/12/2019	23/12/2019	
									A071102	Failure to Interrogate	23/12/2019	23/12/2019	
							F1901	Additional Surgery	A070504	Premature Discharge of Battery	23/12/2019	23/12/2019	
									A071102	Failure to Interrogate	23/12/2019	23/12/2019	
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	23/12/2019	23/12/2019	
									A071102	Failure to Interrogate	23/12/2019	23/12/2019	
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	23/12/2019	23/12/2019	
									A071102	Failure to Interrogate	23/12/2019	23/12/2019	
							F1901	Additional Surgery	A070504	Premature Discharge of Battery	23/12/2019	23/12/2019	
									A071102	Failure to Interrogate	23/12/2019	23/12/2019	
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	23/12/2019	23/12/2019	
									A071102	Failure to Interrogate	23/12/2019	23/12/2019	
					CT2985	Active	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	23/12/2019	23/12/2019	
									A071102	Failure to Interrogate	23/12/2019	23/12/2019	
							F1901	Additional Surgery	A070504	Premature Discharge of Battery	23/12/2019	23/12/2019	
									A071102	Failure to Interrogate	23/12/2019	23/12/2019	
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	23/12/2019	23/12/2019	
									A071102	Failure to Interrogate	23/12/2019	23/12/2019	
					CT3001	Implantable metal	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	23/12/2019	23/12/2019	
									A071102	Failure to Interrogate	23/12/2019	23/12/2019	
							F1901	Additional Surgery	A070504	Premature Discharge of Battery	23/12/2019	23/12/2019	
									A071102	Failure to Interrogate	23/12/2019	23/12/2019	
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	23/12/2019	23/12/2019	
									A071102	Failure to Interrogate	23/12/2019	23/12/2019	
					CT3002	Implantable synthetic polymer	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	23/12/2019	23/12/2019	
									A071102	Failure to Interrogate	23/12/2019	23/12/2019	
							F1901	Additional Surgery	A070504	Premature Discharge of Battery	23/12/2019	23/12/2019	
									A071102	Failure to Interrogate	23/12/2019	23/12/2019	
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	23/12/2019	23/12/2019	
									A071102	Failure to Interrogate	23/12/2019	23/12/2019	
					CT3006	Implantable inorganic material	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	23/12/2019	23/12/2019	
									A071102	Failure to Interrogate	23/12/2019	23/12/2019	

							F1901	Additional Surgery	A070504	Premature Discharge of Battery	23/12/2019	23/12/2019
									A071102	Failure to Interrogate	23/12/2019	23/12/2019
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	23/12/2019	23/12/2019
									A071102	Failure to Interrogate	23/12/2019	23/12/2019
					CT314	Electrically powered	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	23/12/2019	23/12/2019
									A071102	Failure to Interrogate	23/12/2019	23/12/2019
							F1901	Additional Surgery	A070504	Premature Discharge of Battery	23/12/2019	23/12/2019
									A071102	Failure to Interrogate	23/12/2019	23/12/2019
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	23/12/2019	23/12/2019
									A071102	Failure to Interrogate	23/12/2019	23/12/2019
					CT321	Long-term surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	23/12/2019	23/12/2019
									A071102	Failure to Interrogate	23/12/2019	23/12/2019
							F1901	Additional Surgery	A070504	Premature Discharge of Battery	23/12/2019	23/12/2019
									A071102	Failure to Interrogate	23/12/2019	23/12/2019
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	23/12/2019	23/12/2019
									A071102	Failure to Interrogate	23/12/2019	23/12/2019
					CT449	Generators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	23/12/2019	23/12/2019
									A071102	Failure to Interrogate	23/12/2019	23/12/2019
							F1901	Additional Surgery	A070504	Premature Discharge of Battery	23/12/2019	23/12/2019
									A071102	Failure to Interrogate	23/12/2019	23/12/2019
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	23/12/2019	23/12/2019
									A071102	Failure to Interrogate	23/12/2019	23/12/2019
					CT981	Single use	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	23/12/2019	23/12/2019
									A071102	Failure to Interrogate	23/12/2019	23/12/2019
							F1901	Additional Surgery	A070504	Premature Discharge of Battery	23/12/2019	23/12/2019
									A071102	Failure to Interrogate	23/12/2019	23/12/2019
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	23/12/2019	23/12/2019
									A071102	Failure to Interrogate	23/12/2019	23/12/2019
					CT983	Surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	23/12/2019	23/12/2019
									A071102	Failure to Interrogate	23/12/2019	23/12/2019
							F1901	Additional Surgery	A070504	Premature Discharge of Battery	23/12/2019	23/12/2019
									A071102	Failure to Interrogate	23/12/2019	23/12/2019
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	23/12/2019	23/12/2019
									A071102	Failure to Interrogate	23/12/2019	23/12/2019
8326	2019	All other reportable incidents	35852	Single-chamber implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	08/02/2019	
8327	2019	All other reportable incidents	35223	Endocardial/interventricular septal pacing lead	CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070909	Oversensing		
8328	2019	All other reportable incidents	47267	Single-chamber implantable pacemaker, rate-responsive	CT141	Electrodes/Leads and associated devices	F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	12/01/2019	
					CT145	Cardiovascular devices	F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	12/01/2019	
					CT1684	Cardiac pacemakers and associated devices	F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	12/01/2019	
					CT1685	Cardiac pacemakers	F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	12/01/2019	
					CT233	Surgical	F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	12/01/2019	
					CT2406	Implantable	F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	12/01/2019	
					CT2407	Active implantable	F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	12/01/2019	
					CT257	Implantable cardiac pacemakers	F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	12/01/2019	
					CT2727	Active implant devices/ancillaries	F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	12/01/2019	
					CT2750	Implantable cardiac electrical stimulators and associated devices	F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	12/01/2019	
					CT2866	Implantable cardiac electrical stimulators	F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	12/01/2019	
					CT2985	Active	F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	12/01/2019	
					CT3001	Implantable metal	F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	12/01/2019	
					CT3002	Implantable synthetic polymer	F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	12/01/2019	
					CT3006	Implantable inorganic material	F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	12/01/2019	
					CT314	Electrically powered	F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	12/01/2019	
					CT321	Long-term surgical invasive	F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	12/01/2019	
					CT449	Generators and associated devices	F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	12/01/2019	
					CT981	Single use	F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	12/01/2019	
					CT983	Surgical invasive	F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	12/01/2019	
8329	2019	All other reportable incidents	35852	Single-chamber implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	14/02/2019	
8330	2020	All other reportable incidents	47265	Dual-chamber implantable pacemaker, rate-responsive	CT141	Electrodes/Leads and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	12/12/2019	12/12/2019
							F1901	Additional Surgery	A070504	Premature Discharge of Battery	12/12/2019	12/12/2019
					CT145	Cardiovascular devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	12/12/2019	12/12/2019
							F1901	Additional Surgery	A070504	Premature Discharge of Battery	12/12/2019	12/12/2019
					CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	12/12/2019	12/12/2019
							F1901	Additional Surgery	A070504	Premature Discharge of Battery	12/12/2019	12/12/2019
					CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	12/12/2019	12/12/2019
							F1901	Additional Surgery	A070504	Premature Discharge of Battery	12/12/2019	12/12/2019
					CT233	Surgical	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	12/12/2019	12/12/2019
							F1901	Additional Surgery	A070504	Premature Discharge of Battery	12/12/2019	12/12/2019
					CT2406	Implantable	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	12/12/2019	12/12/2019
							F1901	Additional Surgery	A070504	Premature Discharge of Battery	12/12/2019	12/12/2019
					CT2407	Active implantable	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	12/12/2019	12/12/2019
							F1901	Additional Surgery	A070504	Premature Discharge of Battery	12/12/2019	12/12/2019
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	12/12/2019	12/12/2019
							F1901	Additional Surgery	A070504	Premature Discharge of Battery	12/12/2019	12/12/2019
					CT2727	Active implant devices/ancillaries	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	12/12/2019	12/12/2019
							F1901	Additional Surgery	A070504	Premature Discharge of Battery	12/12/2019	12/12/2019
					CT2750	Implantable cardiac electrical stimulators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	12/12/2019	12/12/2019
							F1901	Additional Surgery	A070504	Premature Discharge of Battery	12/12/2019	12/12/2019
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	12/12/2019	12/12/2019
							F1901	Additional Surgery	A070504	Premature Discharge of Battery	12/12/2019	12/12/2019
					CT2985	Active	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	12/12/2019	12/12/2019
							F1901	Additional Surgery	A070504	Premature Discharge of Battery	12/12/2019	12/12/2019
					CT3001	Implantable metal	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	12/12/2019	12/12/2019
							F1901	Additional Surgery	A070504	Premature Discharge of Battery	12/12/2019	12/12/2019
					CT3002	Implantable synthetic polymer	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	12/12/2019	12/12/2019
							F1901	Additional Surgery	A070504	Premature Discharge of Battery	12/12/2019	12/12/2019
					CT3006	Implantable inorganic material	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	12/12/2019	12/12/2019

							F1901	Additional Surgery	A070504	Premature Discharge of Battery	12/12/2019	12/12/2019
					CT314	Electrically powered	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	12/12/2019	12/12/2019
							F1901	Additional Surgery	A070504	Premature Discharge of Battery	12/12/2019	12/12/2019
					CT321	Long-term surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	12/12/2019	12/12/2019
							F1901	Additional Surgery	A070504	Premature Discharge of Battery	12/12/2019	12/12/2019
					CT449	Generators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	12/12/2019	12/12/2019
							F1901	Additional Surgery	A070504	Premature Discharge of Battery	12/12/2019	12/12/2019
					CT981	Single use	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	12/12/2019	12/12/2019
							F1901	Additional Surgery	A070504	Premature Discharge of Battery	12/12/2019	12/12/2019
					CT983	Surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	12/12/2019	12/12/2019
							F1901	Additional Surgery	A070504	Premature Discharge of Battery	12/12/2019	12/12/2019
8331	2019	All other reportable incidents	35852	Single-chamber implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E2104	Electric Shock	A070909	Oversensing		
									A071302	Inappropriate/Inadequate Shock/Stimulation		
							E2403	No Clinical Signs, Symptoms or Conditions	A070909	Oversensing		
									A071302	Inappropriate/Inadequate Shock/Stimulation		
							F26	No Health Consequences or Impact	A070909	Oversensing		
									A071302	Inappropriate/Inadequate Shock/Stimulation		
8332	2019	All other reportable incidents	37265	Dual-chamber implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	21/01/2020	21/01/2020
									A090201	Device Displays Incorrect Message	21/01/2020	21/01/2020
8333	2019	All other reportable incidents	35852	Single-chamber implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A070909	Oversensing		
8334	2020	Death	17882	Physiologic-monitoring defibrillation system	CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A071301	Failure to Deliver Shock/Stimulation		
							F02	Death	A071301	Failure to Deliver Shock/Stimulation		
8335	2019	All other reportable incidents	47265	Dual-chamber implantable pacemaker, rate-responsive	CT141	Electrodes/Leads and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	08/04/2020	08/04/2020
							F12	Serious Injury/ Illness/ Impairment	A070504	Premature Discharge of Battery	08/04/2020	08/04/2020
							F19	Surgical Intervention	A070504	Premature Discharge of Battery	08/04/2020	08/04/2020
					CT145	Cardiovascular devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	08/04/2020	08/04/2020
							F12	Serious Injury/ Illness/ Impairment	A070504	Premature Discharge of Battery	08/04/2020	08/04/2020
							F19	Surgical Intervention	A070504	Premature Discharge of Battery	08/04/2020	08/04/2020
					CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	08/04/2020	08/04/2020
							F12	Serious Injury/ Illness/ Impairment	A070504	Premature Discharge of Battery	08/04/2020	08/04/2020
							F19	Surgical Intervention	A070504	Premature Discharge of Battery	08/04/2020	08/04/2020
					CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	08/04/2020	08/04/2020
							F12	Serious Injury/ Illness/ Impairment	A070504	Premature Discharge of Battery	08/04/2020	08/04/2020
							F19	Surgical Intervention	A070504	Premature Discharge of Battery	08/04/2020	08/04/2020
					CT233	Surgical	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	08/04/2020	08/04/2020
							F12	Serious Injury/ Illness/ Impairment	A070504	Premature Discharge of Battery	08/04/2020	08/04/2020
							F19	Surgical Intervention	A070504	Premature Discharge of Battery	08/04/2020	08/04/2020
					CT2406	Implantable	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	08/04/2020	08/04/2020
							F12	Serious Injury/ Illness/ Impairment	A070504	Premature Discharge of Battery	08/04/2020	08/04/2020
							F19	Surgical Intervention	A070504	Premature Discharge of Battery	08/04/2020	08/04/2020
					CT2407	Active implantable	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	08/04/2020	08/04/2020
							F12	Serious Injury/ Illness/ Impairment	A070504	Premature Discharge of Battery	08/04/2020	08/04/2020
							F19	Surgical Intervention	A070504	Premature Discharge of Battery	08/04/2020	08/04/2020
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	08/04/2020	08/04/2020
							F12	Serious Injury/ Illness/ Impairment	A070504	Premature Discharge of Battery	08/04/2020	08/04/2020
							F19	Surgical Intervention	A070504	Premature Discharge of Battery	08/04/2020	08/04/2020
					CT2727	Active implant devices/ancillaries	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	08/04/2020	08/04/2020
							F12	Serious Injury/ Illness/ Impairment	A070504	Premature Discharge of Battery	08/04/2020	08/04/2020
							F19	Surgical Intervention	A070504	Premature Discharge of Battery	08/04/2020	08/04/2020
					CT2750	Implantable cardiac electrical stimulators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	08/04/2020	08/04/2020
							F12	Serious Injury/ Illness/ Impairment	A070504	Premature Discharge of Battery	08/04/2020	08/04/2020
							F19	Surgical Intervention	A070504	Premature Discharge of Battery	08/04/2020	08/04/2020
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	08/04/2020	08/04/2020
							F12	Serious Injury/ Illness/ Impairment	A070504	Premature Discharge of Battery	08/04/2020	08/04/2020
							F19	Surgical Intervention	A070504	Premature Discharge of Battery	08/04/2020	08/04/2020
					CT2985	Active	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	08/04/2020	08/04/2020
							F12	Serious Injury/ Illness/ Impairment	A070504	Premature Discharge of Battery	08/04/2020	08/04/2020
							F19	Surgical Intervention	A070504	Premature Discharge of Battery	08/04/2020	08/04/2020
					CT3001	Implantable metal	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	08/04/2020	08/04/2020
							F12	Serious Injury/ Illness/ Impairment	A070504	Premature Discharge of Battery	08/04/2020	08/04/2020
							F19	Surgical Intervention	A070504	Premature Discharge of Battery	08/04/2020	08/04/2020
					CT3002	Implantable synthetic polymer	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	08/04/2020	08/04/2020
							F12	Serious Injury/ Illness/ Impairment	A070504	Premature Discharge of Battery	08/04/2020	08/04/2020
							F19	Surgical Intervention	A070504	Premature Discharge of Battery	08/04/2020	08/04/2020
					CT3006	Implantable inorganic material	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	08/04/2020	08/04/2020
							F12	Serious Injury/ Illness/ Impairment	A070504	Premature Discharge of Battery	08/04/2020	08/04/2020
							F19	Surgical Intervention	A070504	Premature Discharge of Battery	08/04/2020	08/04/2020
					CT314	Electrically powered	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	08/04/2020	08/04/2020
							F12	Serious Injury/ Illness/ Impairment	A070504	Premature Discharge of Battery	08/04/2020	08/04/2020
							F19	Surgical Intervention	A070504	Premature Discharge of Battery	08/04/2020	08/04/2020
					CT321	Long-term surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	08/04/2020	08/04/2020
							F12	Serious Injury/ Illness/ Impairment	A070504	Premature Discharge of Battery	08/04/2020	08/04/2020
							F19	Surgical Intervention	A070504	Premature Discharge of Battery	08/04/2020	08/04/2020
					CT449	Generators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	08/04/2020	08/04/2020
							F12	Serious Injury/ Illness/ Impairment	A070504	Premature Discharge of Battery	08/04/2020	08/04/2020
							F19	Surgical Intervention	A070504	Premature Discharge of Battery	08/04/2020	08/04/2020
					CT981	Single use	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	08/04/2020	08/04/2020
							F12	Serious Injury/ Illness/ Impairment	A070504	Premature Discharge of Battery	08/04/2020	08/04/2020
							F19	Surgical Intervention	A070504	Premature Discharge of Battery	08/04/2020	08/04/2020
					CT983	Surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	08/04/2020	08/04/2020
							F12	Serious Injury/ Illness/ Impairment	A070504	Premature Discharge of Battery	08/04/2020	08/04/2020
							F19	Surgical Intervention	A070504	Premature Discharge of Battery	08/04/2020	08/04/2020
8336	2019	All other reportable incidents	35852	Single-chamber implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A070909	Oversensing		
8337	2020	All other reportable incidents	35852	Single-chamber implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A071302	Inappropriate/Inadequate Shock/Stimulation		
							F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A071302	Inappropriate/Inadequate Shock/Stimulation		
							F26	No Health Consequences or Impact	A071302	Inappropriate/Inadequate Shock/Stimulation		
8338	2020	All other reportable incidents	37265	Dual-chamber implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	06/12/2019	06/12/2019
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	06/12/2019	06/12/2019

8339	2020	All other reportable incidents	37265	Dual-chamber implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A070909	Oversensing	06/01/2021	06/01/2021
							F26	No Health Consequences or Impact	A070909	Oversensing	06/01/2021	06/01/2021
8340	2020	All other reportable incidents	36241	Pacemaker/defibrillator lead	CT1684	Cardiac pacemakers and associated devices	F1905	Device Revision or Replacement	A072201	High Impedance	11/07/2019	11/07/2019
8341	2019	All other reportable incidents	36241	Pacemaker/defibrillator lead	CT1684	Cardiac pacemakers and associated devices	E0623	Unspecified Heart Problem	A04	Material Integrity Problem		
8342	2020	All other reportable incidents	47265	Dual-chamber implantable pacemaker, rate-responsive	CT141	Electrodes/Leads and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	12/11/2019	12/11/2019
							F19	Surgical Intervention	A070504	Premature Discharge of Battery	12/11/2019	12/11/2019
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	12/11/2019	12/11/2019
					CT145	Cardiovascular devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	12/11/2019	12/11/2019
							F19	Surgical Intervention	A070504	Premature Discharge of Battery	12/11/2019	12/11/2019
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	12/11/2019	12/11/2019
					CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	12/11/2019	12/11/2019
							F19	Surgical Intervention	A070504	Premature Discharge of Battery	12/11/2019	12/11/2019
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	12/11/2019	12/11/2019
					CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	12/11/2019	12/11/2019
							F19	Surgical Intervention	A070504	Premature Discharge of Battery	12/11/2019	12/11/2019
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	12/11/2019	12/11/2019
					CT233	Surgical	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	12/11/2019	12/11/2019
							F19	Surgical Intervention	A070504	Premature Discharge of Battery	12/11/2019	12/11/2019
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	12/11/2019	12/11/2019
					CT2406	Implantable	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	12/11/2019	12/11/2019
							F19	Surgical Intervention	A070504	Premature Discharge of Battery	12/11/2019	12/11/2019
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	12/11/2019	12/11/2019
					CT2407	Active implantable	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	12/11/2019	12/11/2019
							F19	Surgical Intervention	A070504	Premature Discharge of Battery	12/11/2019	12/11/2019
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	12/11/2019	12/11/2019
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	12/11/2019	12/11/2019
							F19	Surgical Intervention	A070504	Premature Discharge of Battery	12/11/2019	12/11/2019
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	12/11/2019	12/11/2019
					CT2727	Active implant devices/ancillaries	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	12/11/2019	12/11/2019
							F19	Surgical Intervention	A070504	Premature Discharge of Battery	12/11/2019	12/11/2019
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	12/11/2019	12/11/2019
					CT2750	Implantable cardiac electrical stimulators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	12/11/2019	12/11/2019
							F19	Surgical Intervention	A070504	Premature Discharge of Battery	12/11/2019	12/11/2019
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	12/11/2019	12/11/2019
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	12/11/2019	12/11/2019
							F19	Surgical Intervention	A070504	Premature Discharge of Battery	12/11/2019	12/11/2019
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	12/11/2019	12/11/2019
					CT2985	Active	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	12/11/2019	12/11/2019
							F19	Surgical Intervention	A070504	Premature Discharge of Battery	12/11/2019	12/11/2019
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	12/11/2019	12/11/2019
					CT3001	Implantable metal	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	12/11/2019	12/11/2019
							F19	Surgical Intervention	A070504	Premature Discharge of Battery	12/11/2019	12/11/2019
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	12/11/2019	12/11/2019
					CT3002	Implantable synthetic polymer	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	12/11/2019	12/11/2019
							F19	Surgical Intervention	A070504	Premature Discharge of Battery	12/11/2019	12/11/2019
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	12/11/2019	12/11/2019
					CT3006	Implantable inorganic material	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	12/11/2019	12/11/2019
							F19	Surgical Intervention	A070504	Premature Discharge of Battery	12/11/2019	12/11/2019
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	12/11/2019	12/11/2019
					CT314	Electrically powered	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	12/11/2019	12/11/2019
							F19	Surgical Intervention	A070504	Premature Discharge of Battery	12/11/2019	12/11/2019
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	12/11/2019	12/11/2019
					CT321	Long-term surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	12/11/2019	12/11/2019
							F19	Surgical Intervention	A070504	Premature Discharge of Battery	12/11/2019	12/11/2019
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	12/11/2019	12/11/2019
					CT449	Generators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	12/11/2019	12/11/2019
							F19	Surgical Intervention	A070504	Premature Discharge of Battery	12/11/2019	12/11/2019
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	12/11/2019	12/11/2019
					CT981	Single use	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	12/11/2019	12/11/2019
							F19	Surgical Intervention	A070504	Premature Discharge of Battery	12/11/2019	12/11/2019
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	12/11/2019	12/11/2019
					CT983	Surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	12/11/2019	12/11/2019
							F19	Surgical Intervention	A070504	Premature Discharge of Battery	12/11/2019	12/11/2019
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	12/11/2019	12/11/2019
8343	2020	All other reportable incidents	37265	Dual-chamber implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A072201	High Impedance		
8344	2020	All other reportable incidents	36241	Pacemaker/defibrillator lead	CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A072201	High Impedance		
							F24	Insufficient Health Impact Information	A072201	High Impedance		
8345	2019	All other reportable incidents	37265	Dual-chamber implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A072201	High Impedance		
							F26	No Health Consequences or Impact	A072201	High Impedance		
8346	2019	All other reportable incidents	47270	Cardiac resynchronization therapy implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A0709	Device Sensing Problem		
									A070909	Oversensing		
									A0712	Pacing Problem		
									A071209	Extracardiac Stimulation		
8347	2019	All other reportable incidents	36241	Pacemaker/defibrillator lead	CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A072201	High Impedance		
							F26	No Health Consequences or Impact	A072201	High Impedance		
8348	2019	All other reportable incidents	37265	Dual-chamber implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	03/01/2020	03/01/2020
									A090201	Device Displays Incorrect Message	03/01/2020	03/01/2020
8349	2019	All other reportable incidents	47270	Cardiac resynchronization therapy implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	25/02/2019	
8350	2019	All other reportable incidents	47267	Single-chamber implantable pacemaker, rate-responsive	CT141	Electrodes/Leads and associated devices	F1905	Device Revision or Replacement	A071102	Failure to Interrogate	04/02/2019	
					CT145	Cardiovascular devices	F1905	Device Revision or Replacement	A071102	Failure to Interrogate	04/02/2019	
					CT1684	Cardiac pacemakers and associated devices	F1905	Device Revision or Replacement	A071102	Failure to Interrogate	04/02/2019	
					CT1685	Cardiac pacemakers	F1905	Device Revision or Replacement	A071102	Failure to Interrogate	04/02/2019	
					CT233	Surgical	F1905	Device Revision or Replacement	A071102	Failure to Interrogate	04/02/2019	
					CT2406	Implantable	F1905	Device Revision or Replacement	A071102	Failure to Interrogate	04/02/2019	
					CT2407	Active implantable	F1905	Device Revision or Replacement	A071102	Failure to Interrogate	04/02/2019	
					CT257	Implantable cardiac pacemakers	F1905	Device Revision or Replacement	A071102	Failure to Interrogate	04/02/2019	
					CT2727	Active implant devices/ancillaries	F1905	Device Revision or Replacement	A071102	Failure to Interrogate	04/02/2019	

					CT2750	Implantable cardiac electrical stimulators and associated devices	F1905	Device Revision or Replacement	A071102	Failure to Interrogate	04/02/2019		
					CT2866	Implantable cardiac electrical stimulators	F1905	Device Revision or Replacement	A071102	Failure to Interrogate	04/02/2019		
					CT2985	Active	F1905	Device Revision or Replacement	A071102	Failure to Interrogate	04/02/2019		
					CT3001	Implantable metal	F1905	Device Revision or Replacement	A071102	Failure to Interrogate	04/02/2019		
					CT3002	Implantable synthetic polymer	F1905	Device Revision or Replacement	A071102	Failure to Interrogate	04/02/2019		
					CT3006	Implantable inorganic material	F1905	Device Revision or Replacement	A071102	Failure to Interrogate	04/02/2019		
					CT314	Electrically powered	F1905	Device Revision or Replacement	A071102	Failure to Interrogate	04/02/2019		
					CT321	Long-term surgical invasive	F1905	Device Revision or Replacement	A071102	Failure to Interrogate	04/02/2019		
					CT449	Generators and associated devices	F1905	Device Revision or Replacement	A071102	Failure to Interrogate	04/02/2019		
					CT981	Single use	F1905	Device Revision or Replacement	A071102	Failure to Interrogate	04/02/2019		
					CT983	Surgical invasive	F1905	Device Revision or Replacement	A071102	Failure to Interrogate	04/02/2019		
8351	2019	All other reportable incidents	37265	Dual-chamber implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery			
8352	2019	All other reportable incidents	36241	Pacemaker/defibrillator lead	CT1684	Cardiac pacemakers and associated devices	E1906	Unspecified Infection	A010102	Device Appears to Trigger Rejection	25/01/2019		
							F1905	Device Revision or Replacement	A010102	Device Appears to Trigger Rejection	25/01/2019		
8353	2019	All other reportable incidents	35852	Single-chamber implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E1906	Unspecified Infection	A010102	Device Appears to Trigger Rejection	25/01/2019		
							F1905	Device Revision or Replacement	A010102	Device Appears to Trigger Rejection	25/01/2019		
8354	2020	All other reportable incidents	35223	Endocardial/interventricular septal pacing lead	CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A0401	Break	19/12/2019	19/12/2019	
									A070909	Oversensing	19/12/2019	19/12/2019	
									A0712	Pacing Problem	19/12/2019	19/12/2019	
							F1908	Prolonged Surgery	A0401	Break	19/12/2019	19/12/2019	
									A070909	Oversensing	19/12/2019	19/12/2019	
									A0712	Pacing Problem	19/12/2019	19/12/2019	
8355	2020	All other reportable incidents	47267	Single-chamber implantable pacemaker, rate-responsive	CT141	Electrodes/Leads and associated devices	F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery			
					CT145	Cardiovascular devices	F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery			
					CT1684	Cardiac pacemakers and associated devices	F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery			
					CT1685	Cardiac pacemakers	F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery			
					CT233	Surgical	F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery			
					CT2406	Implantable	F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery			
					CT2407	Active implantable	F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery			
					CT257	Implantable cardiac pacemakers	F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery			
					CT2727	Active implant devices/ancillaries	F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery			
					CT2750	Implantable cardiac electrical stimulators and associated devices	F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery			
					CT2866	Implantable cardiac electrical stimulators	F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery			
					CT2985	Active	F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery			
					CT3001	Implantable metal	F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery			
					CT3002	Implantable synthetic polymer	F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery			
					CT3006	Implantable inorganic material	F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery			
					CT314	Electrically powered	F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery			
					CT321	Long-term surgical invasive	F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery			
					CT449	Generators and associated devices	F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery			
					CT981	Single use	F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery			
					CT983	Surgical invasive	F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery			
8356	2019	All other reportable incidents	35852	Single-chamber implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A071302	Inappropriate/Inadequate Shock/Stimulation			
8357	2019	All other reportable incidents	47265	Dual-chamber implantable pacemaker, rate-responsive	CT141	Electrodes/Leads and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery			
					CT145	Cardiovascular devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery			
					CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery			
					CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery			
					CT233	Surgical	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery			
					CT2406	Implantable	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery			
					CT2407	Active implantable	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery			
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery			
					CT2727	Active implant devices/ancillaries	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery			
					CT2750	Implantable cardiac electrical stimulators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery			
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery			
					CT2985	Active	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery			
					CT3001	Implantable metal	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery			
					CT3002	Implantable synthetic polymer	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery			
					CT3006	Implantable inorganic material	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery			
					CT314	Electrically powered	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery			
					CT321	Long-term surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery			
					CT449	Generators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery			
					CT981	Single use	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery			
					CT983	Surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery			
8358	2019	All other reportable incidents	35852	Single-chamber implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A0711	Interrogation Problem			
8359	2019	All other reportable incidents	47265	Dual-chamber implantable pacemaker, rate-responsive	CT141	Electrodes/Leads and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery			
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery			
					CT145	Cardiovascular devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery			
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery			
					CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery			
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery			
					CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery			
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery			
					CT233	Surgical	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery			
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery			
					CT2406	Implantable	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery			
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery			
					CT2407	Active implantable	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery			
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery			
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery			
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery			
					CT2727	Active implant devices/ancillaries	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery			
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery			
					CT2750	Implantable cardiac electrical stimulators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery			
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery			
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery			

						F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery			
					CT2985	Active	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT3001	Implantable metal	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT3002	Implantable synthetic polymer	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT3006	Implantable inorganic material	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT314	Electrically powered	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT321	Long-term surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT449	Generators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT981	Single use	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT983	Surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
8360	2019	All other reportable incidents	47265	Dual-chamber implantable pacemaker, rate-responsive	CT141	Electrodes/Leads and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	15/01/2020	15/01/2020
							F19	Surgical Intervention	A070504	Premature Discharge of Battery	15/01/2020	15/01/2020
					CT145	Cardiovascular devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	15/01/2020	15/01/2020
							F19	Surgical Intervention	A070504	Premature Discharge of Battery	15/01/2020	15/01/2020
					CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	15/01/2020	15/01/2020
							F19	Surgical Intervention	A070504	Premature Discharge of Battery	15/01/2020	15/01/2020
					CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	15/01/2020	15/01/2020
							F19	Surgical Intervention	A070504	Premature Discharge of Battery	15/01/2020	15/01/2020
					CT233	Surgical	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	15/01/2020	15/01/2020
							F19	Surgical Intervention	A070504	Premature Discharge of Battery	15/01/2020	15/01/2020
					CT2406	Implantable	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	15/01/2020	15/01/2020
							F19	Surgical Intervention	A070504	Premature Discharge of Battery	15/01/2020	15/01/2020
					CT2407	Active implantable	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	15/01/2020	15/01/2020
							F19	Surgical Intervention	A070504	Premature Discharge of Battery	15/01/2020	15/01/2020
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	15/01/2020	15/01/2020
							F19	Surgical Intervention	A070504	Premature Discharge of Battery	15/01/2020	15/01/2020
					CT2727	Active implant devices/ancillaries	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	15/01/2020	15/01/2020
							F19	Surgical Intervention	A070504	Premature Discharge of Battery	15/01/2020	15/01/2020
					CT2750	Implantable cardiac electrical stimulators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	15/01/2020	15/01/2020
							F19	Surgical Intervention	A070504	Premature Discharge of Battery	15/01/2020	15/01/2020
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	15/01/2020	15/01/2020
							F19	Surgical Intervention	A070504	Premature Discharge of Battery	15/01/2020	15/01/2020
					CT2985	Active	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	15/01/2020	15/01/2020
							F19	Surgical Intervention	A070504	Premature Discharge of Battery	15/01/2020	15/01/2020
					CT3001	Implantable metal	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	15/01/2020	15/01/2020
							F19	Surgical Intervention	A070504	Premature Discharge of Battery	15/01/2020	15/01/2020
					CT3002	Implantable synthetic polymer	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	15/01/2020	15/01/2020
							F19	Surgical Intervention	A070504	Premature Discharge of Battery	15/01/2020	15/01/2020
					CT3006	Implantable inorganic material	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	15/01/2020	15/01/2020
							F19	Surgical Intervention	A070504	Premature Discharge of Battery	15/01/2020	15/01/2020
					CT314	Electrically powered	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	15/01/2020	15/01/2020
							F19	Surgical Intervention	A070504	Premature Discharge of Battery	15/01/2020	15/01/2020
					CT321	Long-term surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	15/01/2020	15/01/2020
							F19	Surgical Intervention	A070504	Premature Discharge of Battery	15/01/2020	15/01/2020
					CT449	Generators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	15/01/2020	15/01/2020
							F19	Surgical Intervention	A070504	Premature Discharge of Battery	15/01/2020	15/01/2020
					CT981	Single use	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	15/01/2020	15/01/2020
							F19	Surgical Intervention	A070504	Premature Discharge of Battery	15/01/2020	15/01/2020
					CT983	Surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	15/01/2020	15/01/2020
							F19	Surgical Intervention	A070504	Premature Discharge of Battery	15/01/2020	15/01/2020
8361	2019	All other reportable incidents	36241	Pacemaker/defibrillator lead	CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A072201	High Impedance		
							F26	No Health Consequences or Impact	A072201	High Impedance		
8362	2019	All other reportable incidents	37265	Dual-chamber implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	09/12/2019	09/12/2019
8363	2019	All other reportable incidents	47265	Dual-chamber implantable pacemaker, rate-responsive	CT141	Electrodes/Leads and associated devices	F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	04/12/2019	04/12/2019
					CT145	Cardiovascular devices	F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	04/12/2019	04/12/2019
					CT1684	Cardiac pacemakers and associated devices	F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	04/12/2019	04/12/2019
					CT1685	Cardiac pacemakers	F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	04/12/2019	04/12/2019
					CT233	Surgical	F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	04/12/2019	04/12/2019
					CT2406	Implantable	F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	04/12/2019	04/12/2019
					CT2407	Active implantable	F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	04/12/2019	04/12/2019
					CT257	Implantable cardiac pacemakers	F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	04/12/2019	04/12/2019
					CT2727	Active implant devices/ancillaries	F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	04/12/2019	04/12/2019
					CT2750	Implantable cardiac electrical stimulators and associated devices	F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	04/12/2019	04/12/2019
					CT2866	Implantable cardiac electrical stimulators	F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	04/12/2019	04/12/2019
					CT2985	Active	F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	04/12/2019	04/12/2019
					CT3001	Implantable metal	F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	04/12/2019	04/12/2019
					CT3002	Implantable synthetic polymer	F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	04/12/2019	04/12/2019
					CT3006	Implantable inorganic material	F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	04/12/2019	04/12/2019
					CT314	Electrically powered	F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	04/12/2019	04/12/2019
					CT321	Long-term surgical invasive	F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	04/12/2019	04/12/2019
					CT449	Generators and associated devices	F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	04/12/2019	04/12/2019
					CT981	Single use	F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	04/12/2019	04/12/2019
					CT983	Surgical invasive	F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	04/12/2019	04/12/2019
8364	2019	All other reportable incidents	47265	Dual-chamber implantable pacemaker, rate-responsive	CT141	Electrodes/Leads and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	20/01/2020	20/01/2020
							F1903	Device Explantation	A070504	Premature Discharge of Battery	20/01/2020	20/01/2020
					CT145	Cardiovascular devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	20/01/2020	20/01/2020
							F1903	Device Explantation	A070504	Premature Discharge of Battery	20/01/2020	20/01/2020
					CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	20/01/2020	20/01/2020

							F1903	Device Explantation	A070504	Premature Discharge of Battery	20/01/2020	20/01/2020
					CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	20/01/2020	20/01/2020
							F1903	Device Explantation	A070504	Premature Discharge of Battery	20/01/2020	20/01/2020
					CT233	Surgical	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	20/01/2020	20/01/2020
							F1903	Device Explantation	A070504	Premature Discharge of Battery	20/01/2020	20/01/2020
					CT2406	Implantable	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	20/01/2020	20/01/2020
							F1903	Device Explantation	A070504	Premature Discharge of Battery	20/01/2020	20/01/2020
					CT2407	Active implantable	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	20/01/2020	20/01/2020
							F1903	Device Explantation	A070504	Premature Discharge of Battery	20/01/2020	20/01/2020
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	20/01/2020	20/01/2020
							F1903	Device Explantation	A070504	Premature Discharge of Battery	20/01/2020	20/01/2020
					CT2727	Active implant devices/ancillaries	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	20/01/2020	20/01/2020
							F1903	Device Explantation	A070504	Premature Discharge of Battery	20/01/2020	20/01/2020
					CT2750	Implantable cardiac electrical stimulators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	20/01/2020	20/01/2020
							F1903	Device Explantation	A070504	Premature Discharge of Battery	20/01/2020	20/01/2020
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	20/01/2020	20/01/2020
							F1903	Device Explantation	A070504	Premature Discharge of Battery	20/01/2020	20/01/2020
					CT2985	Active	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	20/01/2020	20/01/2020
							F1903	Device Explantation	A070504	Premature Discharge of Battery	20/01/2020	20/01/2020
					CT3001	Implantable metal	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	20/01/2020	20/01/2020
							F1903	Device Explantation	A070504	Premature Discharge of Battery	20/01/2020	20/01/2020
					CT3002	Implantable synthetic polymer	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	20/01/2020	20/01/2020
							F1903	Device Explantation	A070504	Premature Discharge of Battery	20/01/2020	20/01/2020
					CT3006	Implantable inorganic material	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	20/01/2020	20/01/2020
							F1903	Device Explantation	A070504	Premature Discharge of Battery	20/01/2020	20/01/2020
					CT314	Electrically powered	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	20/01/2020	20/01/2020
							F1903	Device Explantation	A070504	Premature Discharge of Battery	20/01/2020	20/01/2020
					CT321	Long-term surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	20/01/2020	20/01/2020
							F1903	Device Explantation	A070504	Premature Discharge of Battery	20/01/2020	20/01/2020
					CT449	Generators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	20/01/2020	20/01/2020
							F1903	Device Explantation	A070504	Premature Discharge of Battery	20/01/2020	20/01/2020
					CT981	Single use	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	20/01/2020	20/01/2020
							F1903	Device Explantation	A070504	Premature Discharge of Battery	20/01/2020	20/01/2020
					CT983	Surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	20/01/2020	20/01/2020
							F1903	Device Explantation	A070504	Premature Discharge of Battery	20/01/2020	20/01/2020
8365	2019	All other reportable incidents	47267	Single-chamber implantable pacemaker, rate-responsive	CT141	Electrodes/Leads and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	28/01/2020	28/01/2020
					CT145	Cardiovascular devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	28/01/2020	28/01/2020
					CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	28/01/2020	28/01/2020
					CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	28/01/2020	28/01/2020
					CT233	Surgical	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	28/01/2020	28/01/2020
					CT2406	Implantable	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	28/01/2020	28/01/2020
					CT2407	Active implantable	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	28/01/2020	28/01/2020
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	28/01/2020	28/01/2020
					CT2727	Active implant devices/ancillaries	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	28/01/2020	28/01/2020
					CT2750	Implantable cardiac electrical stimulators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	28/01/2020	28/01/2020
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	28/01/2020	28/01/2020
					CT2985	Active	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	28/01/2020	28/01/2020
					CT3001	Implantable metal	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	28/01/2020	28/01/2020
					CT3002	Implantable synthetic polymer	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	28/01/2020	28/01/2020
					CT3006	Implantable inorganic material	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	28/01/2020	28/01/2020
					CT314	Electrically powered	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	28/01/2020	28/01/2020
					CT321	Long-term surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	28/01/2020	28/01/2020
					CT449	Generators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	28/01/2020	28/01/2020
					CT981	Single use	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	28/01/2020	28/01/2020
					CT983	Surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	28/01/2020	28/01/2020
8366	2019	All other reportable incidents	37265	Dual-chamber implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	09/01/2020	09/01/2020
									A090201	Device Displays Incorrect Message	09/01/2020	09/01/2020
							F19	Surgical Intervention	A070504	Premature Discharge of Battery	09/01/2020	09/01/2020
									A090201	Device Displays Incorrect Message	09/01/2020	09/01/2020
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	09/01/2020	09/01/2020
									A090201	Device Displays Incorrect Message	09/01/2020	09/01/2020
8367	2019	All other reportable incidents	47940	Triple-chamber implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	F1905	Device Revision or Replacement	A0705	Battery Problem	15/02/2019	
8368	2019	All other reportable incidents	36241	Pacemaker/defibrillator lead	CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A0510	Retraction Problem	05/02/2018	
8369	2019	All other reportable incidents	37265	Dual-chamber implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery		
8370	2019	All other reportable incidents	37265	Dual-chamber implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery		
8371	2020	All other reportable incidents	47265	Dual-chamber implantable pacemaker, rate-responsive	CT141	Electrodes/Leads and associated devices	F1905	Device Revision or Replacement	A070102	High Capture Threshold		
									A070504	Premature Discharge of Battery		
					CT145	Cardiovascular devices	F1905	Device Revision or Replacement	A070102	High Capture Threshold		
									A070504	Premature Discharge of Battery		
					CT1684	Cardiac pacemakers and associated devices	F1905	Device Revision or Replacement	A070102	High Capture Threshold		
									A070504	Premature Discharge of Battery		
					CT1685	Cardiac pacemakers	F1905	Device Revision or Replacement	A070102	High Capture Threshold		
									A070504	Premature Discharge of Battery		
					CT233	Surgical	F1905	Device Revision or Replacement	A070102	High Capture Threshold		
									A070504	Premature Discharge of Battery		
					CT2406	Implantable	F1905	Device Revision or Replacement	A070102	High Capture Threshold		
									A070504	Premature Discharge of Battery		
					CT2407	Active implantable	F1905	Device Revision or Replacement	A070102	High Capture Threshold		
									A070504	Premature Discharge of Battery		
					CT257	Implantable cardiac pacemakers	F1905	Device Revision or Replacement	A070102	High Capture Threshold		
									A070504	Premature Discharge of Battery		
					CT2727	Active implant devices/ancillaries	F1905	Device Revision or Replacement	A070102	High Capture Threshold		
									A070504	Premature Discharge of Battery		
					CT2750	Implantable cardiac electrical stimulators and associated devices	F1905	Device Revision or Replacement	A070102	High Capture Threshold		
									A070504	Premature Discharge of Battery		
					CT2866	Implantable cardiac electrical stimulators	F1905	Device Revision or Replacement	A070102	High Capture Threshold		
									A070504	Premature Discharge of Battery		

					CT2985	Active	F1905	Device Revision or Replacement	A070102	High Capture Threshold		
									A070504	Premature Discharge of Battery		
					CT3001	Implantable metal	F1905	Device Revision or Replacement	A070102	High Capture Threshold		
									A070504	Premature Discharge of Battery		
					CT3002	Implantable synthetic polymer	F1905	Device Revision or Replacement	A070102	High Capture Threshold		
									A070504	Premature Discharge of Battery		
					CT3006	Implantable inorganic material	F1905	Device Revision or Replacement	A070102	High Capture Threshold		
									A070504	Premature Discharge of Battery		
					CT314	Electrically powered	F1905	Device Revision or Replacement	A070102	High Capture Threshold		
									A070504	Premature Discharge of Battery		
					CT321	Long-term surgical invasive	F1905	Device Revision or Replacement	A070102	High Capture Threshold		
									A070504	Premature Discharge of Battery		
					CT449	Generators and associated devices	F1905	Device Revision or Replacement	A070102	High Capture Threshold		
									A070504	Premature Discharge of Battery		
					CT981	Single use	F1905	Device Revision or Replacement	A070102	High Capture Threshold		
									A070504	Premature Discharge of Battery		
					CT983	Surgical invasive	F1905	Device Revision or Replacement	A070102	High Capture Threshold		
									A070504	Premature Discharge of Battery		
8372	2020	All other reportable incidents	37265	Dual-chamber implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	F1905	Device Revision or Replacement	A070909	Oversensing	11/12/2019	11/12/2019
8373	2019	All other reportable incidents	47270	Cardiac resynchronization therapy implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	20/11/2019	20/11/2019
							F1901	Additional Surgery	A070504	Premature Discharge of Battery	20/11/2019	20/11/2019
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	20/11/2019	20/11/2019
8374	2019	All other reportable incidents	47270	Cardiac resynchronization therapy implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	20/11/2019	20/11/2019
									A071102	Failure to Interrogate	20/11/2019	20/11/2019
							F1901	Additional Surgery	A070504	Premature Discharge of Battery	20/11/2019	20/11/2019
									A071102	Failure to Interrogate	20/11/2019	20/11/2019
8375	2019	All other reportable incidents	47265	Dual-chamber implantable pacemaker, rate-responsive	CT141	Electrodes/Leads and associated devices	F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery		
					CT145	Cardiovascular devices	F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery		
					CT1684	Cardiac pacemakers and associated devices	F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery		
					CT1685	Cardiac pacemakers	F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery		
					CT233	Surgical	F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery		
					CT2406	Implantable	F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery		
					CT2407	Active implantable	F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery		
					CT257	Implantable cardiac pacemakers	F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery		
					CT2727	Active implant devices/ancillaries	F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery		
					CT2750	Implantable cardiac electrical stimulators and associated devices	F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery		
					CT2866	Implantable cardiac electrical stimulators	F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery		
					CT2985	Active	F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery		
					CT3001	Implantable metal	F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery		
					CT3002	Implantable synthetic polymer	F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery		
					CT3006	Implantable inorganic material	F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery		
					CT314	Electrically powered	F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery		
					CT321	Long-term surgical invasive	F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery		
					CT449	Generators and associated devices	F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery		
					CT981	Single use	F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery		
					CT983	Surgical invasive	F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery		
8376	2019	All other reportable incidents	37265	Dual-chamber implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	F1905	Device Revision or Replacement	A07	Electrical /Electronic Property Problem	14/02/2019	
8377	2019	All other reportable incidents	47265	Dual-chamber implantable pacemaker, rate-responsive	CT141	Electrodes/Leads and associated devices	F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	13/02/2019	
					CT145	Cardiovascular devices	F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	13/02/2019	
					CT1684	Cardiac pacemakers and associated devices	F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	13/02/2019	
					CT1685	Cardiac pacemakers	F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	13/02/2019	
					CT233	Surgical	F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	13/02/2019	
					CT2406	Implantable	F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	13/02/2019	
					CT2407	Active implantable	F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	13/02/2019	
					CT257	Implantable cardiac pacemakers	F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	13/02/2019	
					CT2727	Active implant devices/ancillaries	F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	13/02/2019	
					CT2750	Implantable cardiac electrical stimulators and associated devices	F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	13/02/2019	
					CT2866	Implantable cardiac electrical stimulators	F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	13/02/2019	
					CT2985	Active	F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	13/02/2019	
					CT3001	Implantable metal	F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	13/02/2019	
					CT3002	Implantable synthetic polymer	F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	13/02/2019	
					CT3006	Implantable inorganic material	F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	13/02/2019	
					CT314	Electrically powered	F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	13/02/2019	
					CT321	Long-term surgical invasive	F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	13/02/2019	
					CT449	Generators and associated devices	F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	13/02/2019	
					CT981	Single use	F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	13/02/2019	
					CT983	Surgical invasive	F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	13/02/2019	
8378	2020	All other reportable incidents	47270	Cardiac resynchronization therapy implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	20/12/2019	20/12/2019
							F1901	Additional Surgery	A070504	Premature Discharge of Battery	20/12/2019	20/12/2019
8379	2020	All other reportable incidents	35852	Single-chamber implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	24/09/2019	24/09/2019
							F1901	Additional Surgery	A070504	Premature Discharge of Battery	24/09/2019	24/09/2019
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	24/09/2019	24/09/2019
8380	2019	All other reportable incidents	47265	Dual-chamber implantable pacemaker, rate-responsive	CT141	Electrodes/Leads and associated devices	F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery		
					CT145	Cardiovascular devices	F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery		
					CT1684	Cardiac pacemakers and associated devices	F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery		
					CT1685	Cardiac pacemakers	F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery		
					CT233	Surgical	F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery		
					CT2406	Implantable	F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery		
					CT2407	Active implantable	F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery		
					CT257	Implantable cardiac pacemakers	F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery		
					CT2727	Active implant devices/ancillaries	F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery		
					CT2750	Implantable cardiac electrical stimulators and associated devices	F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery		
					CT2866	Implantable cardiac electrical stimulators	F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery		
					CT2985	Active	F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery		
					CT3001	Implantable metal	F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery		

					CT3002	Implantable synthetic polymer	F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery			
					CT3006	Implantable inorganic material	F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery			
					CT314	Electrically powered	F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery			
					CT321	Long-term surgical invasive	F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery			
					CT449	Generators and associated devices	F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery			
					CT981	Single use	F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery			
					CT983	Surgical invasive	F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery			
8381	2020	All other reportable incidents	35852	Single-chamber implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	26/11/2019	26/11/2019	
							F1901	Additional Surgery	A070504	Premature Discharge of Battery	26/11/2019	26/11/2019	
8382	2020	All other reportable incidents	47270	Cardiac resynchronization therapy implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A071301	Failure to Deliver Shock/Stimulation			
							F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A071301	Failure to Deliver Shock/Stimulation			
							F23	Unexpected Medical Intervention	A071301	Failure to Deliver Shock/Stimulation			
8383	2020	All other reportable incidents	35852	Single-chamber implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A070603	Failure to Charge	09/12/2019	09/12/2019	
							F1901	Additional Surgery	A070603	Failure to Charge	09/12/2019	09/12/2019	
							F1905	Device Revision or Replacement	A070603	Failure to Charge	09/12/2019	09/12/2019	
8384	2019	All other reportable incidents	37265	Dual-chamber implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E2401	Insufficient Information	A070504	Premature Discharge of Battery	01/04/2019	01/04/2019	
									A0908	Incorrect, Inadequate or Imprecise Result or Readings	01/04/2019	01/04/2019	
							E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	01/04/2019	01/04/2019	
									A0908	Incorrect, Inadequate or Imprecise Result or Readings	01/04/2019	01/04/2019	
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	01/04/2019	01/04/2019	
									A0908	Incorrect, Inadequate or Imprecise Result or Readings	01/04/2019	01/04/2019	
							F1901	Additional Surgery	A070504	Premature Discharge of Battery	01/04/2019	01/04/2019	
									A0908	Incorrect, Inadequate or Imprecise Result or Readings	01/04/2019	01/04/2019	
							F1903	Device Explantation	A070504	Premature Discharge of Battery	01/04/2019	01/04/2019	
									A0908	Incorrect, Inadequate or Imprecise Result or Readings	01/04/2019	01/04/2019	
8385	2019	All other reportable incidents	37265	Dual-chamber implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A070910	Undersensing			
8386	2019	All other reportable incidents	37265	Dual-chamber implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery			
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery			
8387	2019	All other reportable incidents	47265	Dual-chamber implantable pacemaker, rate-responsive	CT141	Electrodes/Leads and associated devices	F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery			
					CT145	Cardiovascular devices	F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery			
					CT1684	Cardiac pacemakers and associated devices	F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery			
					CT1685	Cardiac pacemakers	F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery			
					CT233	Surgical	F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery			
					CT2406	Implantable	F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery			
					CT2407	Active implantable	F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery			
					CT257	Implantable cardiac pacemakers	F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery			
					CT2727	Active implant devices/ancillaries	F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery			
					CT2750	Implantable cardiac electrical stimulators and associated devices	F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery			
					CT2866	Implantable cardiac electrical stimulators	F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery			
					CT2985	Active	F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery			
					CT3001	Implantable metal	F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery			
					CT3002	Implantable synthetic polymer	F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery			
					CT3006	Implantable inorganic material	F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery			
					CT314	Electrically powered	F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery			
					CT321	Long-term surgical invasive	F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery			
					CT449	Generators and associated devices	F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery			
					CT981	Single use	F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery			
					CT983	Surgical invasive	F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery			
8388	2019	All other reportable incidents	17882	Physiologic-monitoring defibrillation system	CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem			
8389	2019	All other reportable incidents	36241	Pacemaker/defibrillator lead	CT1684	Cardiac pacemakers and associated devices	F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A070909	Oversensing			
8390	2019	All other reportable incidents	47263	Cardiac resynchronization therapy implantable pacemaker	CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A071102	Failure to Interrogate			
							F26	No Health Consequences or Impact	A071102	Failure to Interrogate			
					CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A071102	Failure to Interrogate			
							F26	No Health Consequences or Impact	A071102	Failure to Interrogate			
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A071102	Failure to Interrogate			
							F26	No Health Consequences or Impact	A071102	Failure to Interrogate			
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A071102	Failure to Interrogate			
							F26	No Health Consequences or Impact	A071102	Failure to Interrogate			
8391	2019	All other reportable incidents	35852	Single-chamber implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A070909	Oversensing	10/02/2021	10/02/2021	
									A071302	Inappropriate/Inadequate Shock/Stimulation	10/02/2021	10/02/2021	
							F08	Hospitalization or Prolonged Hospitalization	A070909	Oversensing	10/02/2021	10/02/2021	
									A071302	Inappropriate/Inadequate Shock/Stimulation	10/02/2021	10/02/2021	
							F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A070909	Oversensing	10/02/2021	10/02/2021	
									A071302	Inappropriate/Inadequate Shock/Stimulation	10/02/2021	10/02/2021	
8392	2019	All other reportable incidents	47270	Cardiac resynchronization therapy implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	29/11/2019	29/11/2019	
							F1901	Additional Surgery	A070504	Premature Discharge of Battery	29/11/2019	29/11/2019	
8393	2019	All other reportable incidents	35223	Endocardial/interventricular septal pacing lead	CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070909	Oversensing			
8394	2019	All other reportable incidents	35223	Endocardial/interventricular septal pacing lead	CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A0401	Break			
									A070909	Oversensing			
							F26	No Health Consequences or Impact	A0401	Break			
									A070909	Oversensing			
8395	2019	All other reportable incidents	35223	Endocardial/interventricular septal pacing lead	CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A0401	Break			
									A070909	Oversensing			
							F26	No Health Consequences or Impact	A0401	Break			
									A070909	Oversensing			
8396	2019	All other reportable incidents	36241	Pacemaker/defibrillator lead	CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A0510	Retraction Problem	05/02/2019		
8397	2019	All other reportable incidents	35223	Endocardial/interventricular septal pacing lead	CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A0701	Capturing Problem	21/11/2019	21/11/2019	
							F1901	Additional Surgery	A0701	Capturing Problem	21/11/2019	21/11/2019	
8398	2019	All other reportable incidents	35852	Single-chamber implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A070909	Oversensing			
8399	2019	All other reportable incidents	35852	Single-chamber implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	11/12/2019	11/12/2019	
							F1901	Additional Surgery	A070504	Premature Discharge of Battery	11/12/2019	11/12/2019	
8400	2020	All other reportable incidents	47265	Dual-chamber implantable pacemaker, rate-	CT141	Electrodes/Leads and associated devices	F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	23/12/2019	23/12/2019	

				responsive	CT145	Cardiovascular devices	F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	23/12/2019	23/12/2019
					CT1684	Cardiac pacemakers and associated devices	F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	23/12/2019	23/12/2019
					CT1685	Cardiac pacemakers	F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	23/12/2019	23/12/2019
					CT233	Surgical	F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	23/12/2019	23/12/2019
					CT2406	Implantable	F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	23/12/2019	23/12/2019
					CT2407	Active implantable	F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	23/12/2019	23/12/2019
					CT257	Implantable cardiac pacemakers	F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	23/12/2019	23/12/2019
					CT2727	Active implant devices/ancillaries	F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	23/12/2019	23/12/2019
					CT2750	Implantable cardiac electrical stimulators and associated devices	F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	23/12/2019	23/12/2019
					CT2866	Implantable cardiac electrical stimulators	F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	23/12/2019	23/12/2019
					CT2985	Active	F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	23/12/2019	23/12/2019
					CT3001	Implantable metal	F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	23/12/2019	23/12/2019
					CT3002	Implantable synthetic polymer	F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	23/12/2019	23/12/2019
					CT3006	Implantable inorganic material	F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	23/12/2019	23/12/2019
					CT314	Electrically powered	F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	23/12/2019	23/12/2019
					CT321	Long-term surgical invasive	F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	23/12/2019	23/12/2019
					CT449	Generators and associated devices	F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	23/12/2019	23/12/2019
					CT981	Single use	F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	23/12/2019	23/12/2019
					CT983	Surgical invasive	F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	23/12/2019	23/12/2019
8401	2019	All other reportable incidents	47265	Dual-chamber implantable pacemaker, rate-responsive	CT141	Electrodes/Leads and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	12/11/2019	12/11/2019
					CT145	Cardiovascular devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	12/11/2019	12/11/2019
					CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	12/11/2019	12/11/2019
					CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	12/11/2019	12/11/2019
					CT233	Surgical	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	12/11/2019	12/11/2019
					CT2406	Implantable	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	12/11/2019	12/11/2019
					CT2407	Active implantable	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	12/11/2019	12/11/2019
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	12/11/2019	12/11/2019
					CT2727	Active implant devices/ancillaries	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	12/11/2019	12/11/2019
					CT2750	Implantable cardiac electrical stimulators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	12/11/2019	12/11/2019
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	12/11/2019	12/11/2019
					CT2985	Active	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	12/11/2019	12/11/2019
					CT3001	Implantable metal	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	12/11/2019	12/11/2019
					CT3002	Implantable synthetic polymer	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	12/11/2019	12/11/2019
					CT3006	Implantable inorganic material	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	12/11/2019	12/11/2019
					CT314	Electrically powered	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	12/11/2019	12/11/2019
					CT321	Long-term surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	12/11/2019	12/11/2019
					CT449	Generators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	12/11/2019	12/11/2019
					CT981	Single use	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	12/11/2019	12/11/2019
					CT983	Surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	12/11/2019	12/11/2019
8402	2019	All other reportable incidents	17882	Physiologic-monitoring defibrillation system	CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070802	Intermittent Loss of Power		
							F05	Delay to Treatment/ Therapy	A070802	Intermittent Loss of Power		
							F26	No Health Consequences or Impact	A070802	Intermittent Loss of Power		
8403	2020	All other reportable incidents	37265	Dual-chamber implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	27/12/2019	27/12/2019
							F1901	Additional Surgery	A070504	Premature Discharge of Battery	27/12/2019	27/12/2019
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	27/12/2019	27/12/2019
8404	2020	All other reportable incidents	35223	Endocardial/interventricular septal pacing lead	CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A0401	Break		
									A070909	Oversensing		
									A0712	Pacing Problem		
							F1908	Prolonged Surgery	A0401	Break		
									A070909	Oversensing		
									A0712	Pacing Problem		
8405	2020	All other reportable incidents	35223	Endocardial/interventricular septal pacing lead	CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A040507	Naturally Worn		
							F23	Unexpected Medical Intervention	A040507	Naturally Worn		
8406	2020	All other reportable incidents	36241	Pacemaker/defibrillator lead	CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A072201	High Impedance		
							F26	No Health Consequences or Impact	A072201	High Impedance		
8407	2020	All other reportable incidents	35852	Single-chamber implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A071302	Inappropriate/Inadequate Shock/Stimulation		
8408	2020	All other reportable incidents	47265	Dual-chamber implantable pacemaker, rate-responsive	CT141	Electrodes/Leads and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT145	Cardiovascular devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT233	Surgical	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT2406	Implantable	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT2407	Active implantable	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT2727	Active implant devices/ancillaries	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT2750	Implantable cardiac electrical stimulators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT2985	Active	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT3001	Implantable metal	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT3002	Implantable synthetic polymer	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT3006	Implantable inorganic material	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		

					CT314	Electrically powered	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT321	Long-term surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT449	Generators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT981	Single use	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT983	Surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
8409	2020	All other reportable incidents	35852	Single-chamber implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A071302	Inappropriate/Inadequate Shock/Stimulation		
8410	2019	All other reportable incidents	35852	Single-chamber implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A070603	Failure to Charge	27/12/2019	27/12/2019
							F1901	Additional Surgery	A070603	Failure to Charge	27/12/2019	27/12/2019
							F1905	Device Revision or Replacement	A070603	Failure to Charge	27/12/2019	27/12/2019
8411	2020	All other reportable incidents	47265	Dual-chamber implantable pacemaker, rate-responsive	CT141	Electrodes/Leads and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070909	Oversensing		
							F26	No Health Consequences or Impact	A070909	Oversensing		
					CT145	Cardiovascular devices	E2403	No Clinical Signs, Symptoms or Conditions	A070909	Oversensing		
							F26	No Health Consequences or Impact	A070909	Oversensing		
					CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070909	Oversensing		
							F26	No Health Consequences or Impact	A070909	Oversensing		
					CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A070909	Oversensing		
							F26	No Health Consequences or Impact	A070909	Oversensing		
					CT233	Surgical	E2403	No Clinical Signs, Symptoms or Conditions	A070909	Oversensing		
							F26	No Health Consequences or Impact	A070909	Oversensing		
					CT2406	Implantable	E2403	No Clinical Signs, Symptoms or Conditions	A070909	Oversensing		
							F26	No Health Consequences or Impact	A070909	Oversensing		
					CT2407	Active implantable	E2403	No Clinical Signs, Symptoms or Conditions	A070909	Oversensing		
							F26	No Health Consequences or Impact	A070909	Oversensing		
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A070909	Oversensing		
							F26	No Health Consequences or Impact	A070909	Oversensing		
					CT2727	Active implant devices/ancillaries	E2403	No Clinical Signs, Symptoms or Conditions	A070909	Oversensing		
							F26	No Health Consequences or Impact	A070909	Oversensing		
					CT2750	Implantable cardiac electrical stimulators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070909	Oversensing		
							F26	No Health Consequences or Impact	A070909	Oversensing		
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A070909	Oversensing		
							F26	No Health Consequences or Impact	A070909	Oversensing		
					CT2985	Active	E2403	No Clinical Signs, Symptoms or Conditions	A070909	Oversensing		
							F26	No Health Consequences or Impact	A070909	Oversensing		
					CT3001	Implantable metal	E2403	No Clinical Signs, Symptoms or Conditions	A070909	Oversensing		
							F26	No Health Consequences or Impact	A070909	Oversensing		
					CT3002	Implantable synthetic polymer	E2403	No Clinical Signs, Symptoms or Conditions	A070909	Oversensing		
							F26	No Health Consequences or Impact	A070909	Oversensing		
					CT3006	Implantable inorganic material	E2403	No Clinical Signs, Symptoms or Conditions	A070909	Oversensing		
							F26	No Health Consequences or Impact	A070909	Oversensing		
					CT314	Electrically powered	E2403	No Clinical Signs, Symptoms or Conditions	A070909	Oversensing		
							F26	No Health Consequences or Impact	A070909	Oversensing		
					CT321	Long-term surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A070909	Oversensing		
							F26	No Health Consequences or Impact	A070909	Oversensing		
					CT449	Generators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070909	Oversensing		
							F26	No Health Consequences or Impact	A070909	Oversensing		
					CT981	Single use	E2403	No Clinical Signs, Symptoms or Conditions	A070909	Oversensing		
							F26	No Health Consequences or Impact	A070909	Oversensing		
					CT983	Surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A070909	Oversensing		
							F26	No Health Consequences or Impact	A070909	Oversensing		
8412	2019	All other reportable incidents	35852	Single-chamber implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	26/11/2019	26/11/2019
							F1901	Additional Surgery	A070504	Premature Discharge of Battery	26/11/2019	26/11/2019
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	26/11/2019	26/11/2019
8413	2019	All other reportable incidents	47267	Single-chamber implantable pacemaker, rate-responsive	CT141	Electrodes/Leads and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070501	Battery Problem: High Impedance		
									A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070501	Battery Problem: High Impedance		
									A070504	Premature Discharge of Battery		
					CT145	Cardiovascular devices	E2403	No Clinical Signs, Symptoms or Conditions	A070501	Battery Problem: High Impedance		
									A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070501	Battery Problem: High Impedance		
									A070504	Premature Discharge of Battery		
					CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070501	Battery Problem: High Impedance		
									A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070501	Battery Problem: High Impedance		
									A070504	Premature Discharge of Battery		
					CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A070501	Battery Problem: High Impedance		
									A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070501	Battery Problem: High Impedance		
									A070504	Premature Discharge of Battery		
					CT233	Surgical	E2403	No Clinical Signs, Symptoms or Conditions	A070501	Battery Problem: High Impedance		
									A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070501	Battery Problem: High Impedance		
									A070504	Premature Discharge of Battery		
					CT2406	Implantable	E2403	No Clinical Signs, Symptoms or Conditions	A070501	Battery Problem: High Impedance		
									A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070501	Battery Problem: High Impedance		
									A070504	Premature Discharge of Battery		
					CT2407	Active implantable	E2403	No Clinical Signs, Symptoms or Conditions	A070501	Battery Problem: High Impedance		
									A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070501	Battery Problem: High Impedance		
									A070504	Premature Discharge of Battery		
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A070501	Battery Problem: High Impedance		
									A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070501	Battery Problem: High Impedance		

									A070504	Premature Discharge of Battery		
					CT2727	Active implant devices/ancillaries	E2403	No Clinical Signs, Symptoms or Conditions	A070501	Battery Problem: High Impedance		
									A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070501	Battery Problem: High Impedance		
									A070504	Premature Discharge of Battery		
					CT2750	Implantable cardiac electrical stimulators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070501	Battery Problem: High Impedance		
									A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070501	Battery Problem: High Impedance		
									A070504	Premature Discharge of Battery		
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A070501	Battery Problem: High Impedance		
									A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070501	Battery Problem: High Impedance		
									A070504	Premature Discharge of Battery		
					CT2985	Active	E2403	No Clinical Signs, Symptoms or Conditions	A070501	Battery Problem: High Impedance		
									A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070501	Battery Problem: High Impedance		
									A070504	Premature Discharge of Battery		
					CT3001	Implantable metal	E2403	No Clinical Signs, Symptoms or Conditions	A070501	Battery Problem: High Impedance		
									A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070501	Battery Problem: High Impedance		
									A070504	Premature Discharge of Battery		
					CT3002	Implantable synthetic polymer	E2403	No Clinical Signs, Symptoms or Conditions	A070501	Battery Problem: High Impedance		
									A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070501	Battery Problem: High Impedance		
									A070504	Premature Discharge of Battery		
					CT3006	Implantable inorganic material	E2403	No Clinical Signs, Symptoms or Conditions	A070501	Battery Problem: High Impedance		
									A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070501	Battery Problem: High Impedance		
									A070504	Premature Discharge of Battery		
					CT314	Electrically powered	E2403	No Clinical Signs, Symptoms or Conditions	A070501	Battery Problem: High Impedance		
									A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070501	Battery Problem: High Impedance		
									A070504	Premature Discharge of Battery		
					CT321	Long-term surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A070501	Battery Problem: High Impedance		
									A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070501	Battery Problem: High Impedance		
									A070504	Premature Discharge of Battery		
					CT449	Generators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070501	Battery Problem: High Impedance		
									A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070501	Battery Problem: High Impedance		
									A070504	Premature Discharge of Battery		
					CT981	Single use	E2403	No Clinical Signs, Symptoms or Conditions	A070501	Battery Problem: High Impedance		
									A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070501	Battery Problem: High Impedance		
									A070504	Premature Discharge of Battery		
					CT983	Surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A070501	Battery Problem: High Impedance		
									A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070501	Battery Problem: High Impedance		
									A070504	Premature Discharge of Battery		
8414	2019	All other reportable incidents	47263	Cardiac resynchronization therapy implantable pacemaker	CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	18/12/2019	18/12/2019
							F1901	Additional Surgery	A070504	Premature Discharge of Battery	18/12/2019	18/12/2019
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	18/12/2019	18/12/2019
					CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	18/12/2019	18/12/2019
							F1901	Additional Surgery	A070504	Premature Discharge of Battery	18/12/2019	18/12/2019
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	18/12/2019	18/12/2019
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	18/12/2019	18/12/2019
							F1901	Additional Surgery	A070504	Premature Discharge of Battery	18/12/2019	18/12/2019
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	18/12/2019	18/12/2019
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	18/12/2019	18/12/2019
							F1901	Additional Surgery	A070504	Premature Discharge of Battery	18/12/2019	18/12/2019
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	18/12/2019	18/12/2019
8415	2019	All other reportable incidents	47263	Cardiac resynchronization therapy implantable pacemaker	CT1684	Cardiac pacemakers and associated devices	F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery		
					CT1685	Cardiac pacemakers	F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery		
					CT257	Implantable cardiac pacemakers	F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery		
					CT2866	Implantable cardiac electrical stimulators	F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery		
8416	2019	All other reportable incidents	47265	Dual-chamber implantable pacemaker, rate-responsive	CT141	Electrodes/Leads and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
					CT145	Cardiovascular devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
					CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
					CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
					CT233	Surgical	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
					CT2406	Implantable	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
					CT2407	Active implantable	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
					CT2727	Active implant devices/ancillaries	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
					CT2750	Implantable cardiac electrical stimulators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
					CT2985	Active	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
					CT3001	Implantable metal	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
					CT3002	Implantable synthetic polymer	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
					CT3006	Implantable inorganic material	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
					CT314	Electrically powered	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
					CT321	Long-term surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
					CT449	Generators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
					CT981	Single use	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
					CT983	Surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
8417	2019	All other reportable incidents	35852	Single-chamber implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E060102	Atrial Fibrillation	A070909	Oversensing		
									A071302	Inappropriate/Inadequate Shock/Stimulation		

							E2403	No Clinical Signs, Symptoms or Conditions	A070909	Oversensing		
									A071302	Inappropriate/Inadequate Shock/Stimulation		
							F26	No Health Consequences or Impact	A070909	Oversensing		
									A071302	Inappropriate/Inadequate Shock/Stimulation		
8418	2020	All other reportable incidents	35852	Single-chamber implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	03/01/2020	03/01/2020
8419	2019	All other reportable incidents	47270	Cardiac resynchronization therapy implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	27/02/2019	
8420	2019	All other reportable incidents	35852	Single-chamber implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A070909	Oversensing		
8421	2019	All other reportable incidents	47265	Dual-chamber implantable pacemaker, rate-responsive	CT141	Electrodes/Leads and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
					CT145	Cardiovascular devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
					CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
					CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
					CT233	Surgical	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
					CT2406	Implantable	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
					CT2407	Active implantable	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
					CT2727	Active implant devices/ancillaries	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
					CT2750	Implantable cardiac electrical stimulators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
					CT2985	Active	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
					CT3001	Implantable metal	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
					CT3002	Implantable synthetic polymer	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
					CT3006	Implantable inorganic material	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
					CT314	Electrically powered	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
					CT321	Long-term surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
					CT449	Generators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
					CT981	Single use	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
					CT983	Surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
8422	2020	All other reportable incidents	35852	Single-chamber implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A070909	Oversensing		
									A071302	Inappropriate/Inadequate Shock/Stimulation		
							F26	No Health Consequences or Impact	A070909	Oversensing		
									A071302	Inappropriate/Inadequate Shock/Stimulation		
8423	2020	All other reportable incidents	37265	Dual-chamber implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A072201	High Impedance		
							F26	No Health Consequences or Impact	A072201	High Impedance		
8424	2020	All other reportable incidents	35852	Single-chamber implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E2401	Insufficient Information	A160601	Premature Elective Replacement Indicator		
							E2403	No Clinical Signs, Symptoms or Conditions	A160601	Premature Elective Replacement Indicator		
							F24	Insufficient Health Impact Information	A160601	Premature Elective Replacement Indicator		
8425	2019	All other reportable incidents	47270	Cardiac resynchronization therapy implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	15/02/2019	
8426	2020	All other reportable incidents	17882	Physiologic-monitoring defibrillation system	CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070908	Failure to Sense		
							F26	No Health Consequences or Impact	A070908	Failure to Sense		
8427	2019	All other reportable incidents	37265	Dual-chamber implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A072201	High Impedance		
8428	2019	All other reportable incidents	36241	Pacemaker/defibrillator lead	CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A072201	High Impedance		
8429	2019	All other reportable incidents	36241	Pacemaker/defibrillator lead	CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070909	Oversensing		
8430	2019	All other reportable incidents	36241	Pacemaker/defibrillator lead	CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070909	Oversensing		
8431	2019	All other reportable incidents	35852	Single-chamber implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A070909	Oversensing		
8432	2019	All other reportable incidents	17882	Physiologic-monitoring defibrillation system	CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A090206	No Display/Image		
8433	2019	All other reportable incidents	36241	Pacemaker/defibrillator lead	CT1684	Cardiac pacemakers and associated devices	F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A0722	Impedance Problem	07/02/2019	
8434	2019	All other reportable incidents	36241	Pacemaker/defibrillator lead	CT1684	Cardiac pacemakers and associated devices	F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A0712	Pacing Problem	07/02/2019	
8435	2019	All other reportable incidents	37265	Dual-chamber implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A070909	Oversensing	22/02/2021	22/02/2021
									A0712	Pacing Problem	22/02/2021	22/02/2021
									A090801	Signal Artifact/Noise	22/02/2021	22/02/2021
							F08	Hospitalization or Prolonged Hospitalization	A070909	Oversensing	22/02/2021	22/02/2021
									A0712	Pacing Problem	22/02/2021	22/02/2021
									A090801	Signal Artifact/Noise	22/02/2021	22/02/2021
							F1905	Device Revision or Replacement	A070909	Oversensing	22/02/2021	22/02/2021
									A0712	Pacing Problem	22/02/2021	22/02/2021
									A090801	Signal Artifact/Noise	22/02/2021	22/02/2021
8436	2019	All other reportable incidents	36241	Pacemaker/defibrillator lead	CT1684	Cardiac pacemakers and associated devices	F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A071302	Inappropriate/Inadequate Shock/Stimulation		
8437	2019	All other reportable incidents	35852	Single-chamber implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A070909	Oversensing		
8438	2020	All other reportable incidents	47265	Dual-chamber implantable pacemaker, rate-responsive	CT141	Electrodes/Leads and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A04	Material Integrity Problem	20/01/2020	20/01/2020
							F1905	Device Revision or Replacement	A04	Material Integrity Problem	20/01/2020	20/01/2020
					CT145	Cardiovascular devices	E2403	No Clinical Signs, Symptoms or Conditions	A04	Material Integrity Problem	20/01/2020	20/01/2020
							F1905	Device Revision or Replacement	A04	Material Integrity Problem	20/01/2020	20/01/2020
					CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A04	Material Integrity Problem	20/01/2020	20/01/2020
							F1905	Device Revision or Replacement	A04	Material Integrity Problem	20/01/2020	20/01/2020
					CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A04	Material Integrity Problem	20/01/2020	20/01/2020
							F1905	Device Revision or Replacement	A04	Material Integrity Problem	20/01/2020	20/01/2020
					CT233	Surgical	E2403	No Clinical Signs, Symptoms or Conditions	A04	Material Integrity Problem	20/01/2020	20/01/2020
							F1905	Device Revision or Replacement	A04	Material Integrity Problem	20/01/2020	20/01/2020
					CT2406	Implantable	E2403	No Clinical Signs, Symptoms or Conditions	A04	Material Integrity Problem	20/01/2020	20/01/2020
							F1905	Device Revision or Replacement	A04	Material Integrity Problem	20/01/2020	20/01/2020
					CT2407	Active implantable	E2403	No Clinical Signs, Symptoms or Conditions	A04	Material Integrity Problem	20/01/2020	20/01/2020
							F1905	Device Revision or Replacement	A04	Material Integrity Problem	20/01/2020	20/01/2020
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A04	Material Integrity Problem	20/01/2020	20/01/2020
							F1905	Device Revision or Replacement	A04	Material Integrity Problem	20/01/2020	20/01/2020
					CT2727	Active implant devices/ancillaries	E2403	No Clinical Signs, Symptoms or Conditions	A04	Material Integrity Problem	20/01/2020	20/01/2020
							F1905	Device Revision or Replacement	A04	Material Integrity Problem	20/01/2020	20/01/2020
					CT2750	Implantable cardiac electrical stimulators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A04	Material Integrity Problem	20/01/2020	20/01/2020
							F1905	Device Revision or Replacement	A04	Material Integrity Problem	20/01/2020	20/01/2020
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A04	Material Integrity Problem	20/01/2020	20/01/2020
							F1905	Device Revision or Replacement	A04	Material Integrity Problem	20/01/2020	20/01/2020
					CT2985	Active	E2403	No Clinical Signs, Symptoms or Conditions	A04	Material Integrity Problem	20/01/2020	20/01/2020
							F1905	Device Revision or Replacement	A04	Material Integrity Problem	20/01/2020	20/01/2020
					CT3001	Implantable metal	E2403	No Clinical Signs, Symptoms or Conditions	A04	Material Integrity Problem	20/01/2020	20/01/2020
							F1905	Device Revision or Replacement	A04	Material Integrity Problem	20/01/2020	20/01/2020
					CT3002	Implantable synthetic polymer	E2403	No Clinical Signs, Symptoms or Conditions	A04	Material Integrity Problem	20/01/2020	20/01/2020

							F1905	Device Revision or Replacement	A04	Material Integrity Problem	20/01/2020	20/01/2020
					CT3006	Implantable inorganic material	E2403	No Clinical Signs, Symptoms or Conditions	A04	Material Integrity Problem	20/01/2020	20/01/2020
							F1905	Device Revision or Replacement	A04	Material Integrity Problem	20/01/2020	20/01/2020
					CT314	Electrically powered	E2403	No Clinical Signs, Symptoms or Conditions	A04	Material Integrity Problem	20/01/2020	20/01/2020
							F1905	Device Revision or Replacement	A04	Material Integrity Problem	20/01/2020	20/01/2020
					CT321	Long-term surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A04	Material Integrity Problem	20/01/2020	20/01/2020
							F1905	Device Revision or Replacement	A04	Material Integrity Problem	20/01/2020	20/01/2020
					CT449	Generators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A04	Material Integrity Problem	20/01/2020	20/01/2020
							F1905	Device Revision or Replacement	A04	Material Integrity Problem	20/01/2020	20/01/2020
					CT981	Single use	E2403	No Clinical Signs, Symptoms or Conditions	A04	Material Integrity Problem	20/01/2020	20/01/2020
							F1905	Device Revision or Replacement	A04	Material Integrity Problem	20/01/2020	20/01/2020
					CT983	Surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A04	Material Integrity Problem	20/01/2020	20/01/2020
							F1905	Device Revision or Replacement	A04	Material Integrity Problem	20/01/2020	20/01/2020
8439	2019	All other reportable incidents	35223	Endocardial/interventricular septal pacing lead	CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A072202	Low Impedance		
8440	2019	All other reportable incidents	35223	Endocardial/interventricular septal pacing lead	CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A072202	Low Impedance		
8441	2019	All other reportable incidents	47270	Cardiac resynchronization therapy implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A070101	Failure to Capture		
8442	2019	All other reportable incidents	36241	Pacemaker/defibrillator lead	CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070101	Failure to Capture	04/02/2019	
8443	2019	All other reportable incidents	35852	Single-chamber implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A071302	Inappropriate/Inadequate Shock/Stimulation		
8444	2019	All other reportable incidents	36241	Pacemaker/defibrillator lead	CT1684	Cardiac pacemakers and associated devices	F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A070909	Oversensing		
8445	2019	All other reportable incidents	35852	Single-chamber implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A070910	Undersensing		
8446	2019	All other reportable incidents	36241	Pacemaker/defibrillator lead	CT1684	Cardiac pacemakers and associated devices	F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A0714	Unintended Electrical Shock		
8447	2019	All other reportable incidents	35852	Single-chamber implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A071302	Inappropriate/Inadequate Shock/Stimulation		
8448	2019	All other reportable incidents	47265	Dual-chamber implantable pacemaker, rate-responsive	CT141	Electrodes/Leads and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A072201	High Impedance		
					CT145	Cardiovascular devices	E2403	No Clinical Signs, Symptoms or Conditions	A072201	High Impedance		
					CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A072201	High Impedance		
					CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A072201	High Impedance		
					CT233	Surgical	E2403	No Clinical Signs, Symptoms or Conditions	A072201	High Impedance		
					CT2406	Implantable	E2403	No Clinical Signs, Symptoms or Conditions	A072201	High Impedance		
					CT2407	Active implantable	E2403	No Clinical Signs, Symptoms or Conditions	A072201	High Impedance		
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A072201	High Impedance		
					CT2727	Active implant devices/ancillaries	E2403	No Clinical Signs, Symptoms or Conditions	A072201	High Impedance		
					CT2750	Implantable cardiac electrical stimulators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A072201	High Impedance		
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A072201	High Impedance		
					CT2985	Active	E2403	No Clinical Signs, Symptoms or Conditions	A072201	High Impedance		
					CT3001	Implantable metal	E2403	No Clinical Signs, Symptoms or Conditions	A072201	High Impedance		
					CT3002	Implantable synthetic polymer	E2403	No Clinical Signs, Symptoms or Conditions	A072201	High Impedance		
					CT3006	Implantable inorganic material	E2403	No Clinical Signs, Symptoms or Conditions	A072201	High Impedance		
					CT314	Electrically powered	E2403	No Clinical Signs, Symptoms or Conditions	A072201	High Impedance		
					CT321	Long-term surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A072201	High Impedance		
					CT449	Generators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A072201	High Impedance		
					CT981	Single use	E2403	No Clinical Signs, Symptoms or Conditions	A072201	High Impedance		
					CT983	Surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A072201	High Impedance		
8449	2019	All other reportable incidents	35852	Single-chamber implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A0709	Device Sensing Problem		
8450	2020	All other reportable incidents	47270	Cardiac resynchronization therapy implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	08/01/2020	08/01/2020
							F1901	Additional Surgery	A070504	Premature Discharge of Battery	08/01/2020	08/01/2020
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	08/01/2020	08/01/2020
8451	2020	All other reportable incidents	47263	Cardiac resynchronization therapy implantable pacemaker	CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A072201	High Impedance		
							F26	No Health Consequences or Impact	A072201	High Impedance		
					CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A072201	High Impedance		
							F26	No Health Consequences or Impact	A072201	High Impedance		
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A072201	High Impedance		
							F26	No Health Consequences or Impact	A072201	High Impedance		
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A072201	High Impedance		
							F26	No Health Consequences or Impact	A072201	High Impedance		
8452	2020	All other reportable incidents	37265	Dual-chamber implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A04	Material Integrity Problem		
8453	2020	All other reportable incidents	17882	Physiologic-monitoring defibrillation system	CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070908	Failure to Sense		
							F26	No Health Consequences or Impact	A070908	Failure to Sense		
8454	2019	All other reportable incidents	47265	Dual-chamber implantable pacemaker, rate-responsive	CT141	Electrodes/Leads and associated devices	F1905	Device Revision or Replacement	A	Annex A	01/02/2019	
					CT145	Cardiovascular devices	F1905	Device Revision or Replacement	A	Annex A	01/02/2019	
					CT1684	Cardiac pacemakers and associated devices	F1905	Device Revision or Replacement	A	Annex A	01/02/2019	
					CT1685	Cardiac pacemakers	F1905	Device Revision or Replacement	A	Annex A	01/02/2019	
					CT233	Surgical	F1905	Device Revision or Replacement	A	Annex A	01/02/2019	
					CT2406	Implantable	F1905	Device Revision or Replacement	A	Annex A	01/02/2019	
					CT2407	Active implantable	F1905	Device Revision or Replacement	A	Annex A	01/02/2019	
					CT257	Implantable cardiac pacemakers	F1905	Device Revision or Replacement	A	Annex A	01/02/2019	
					CT2727	Active implant devices/ancillaries	F1905	Device Revision or Replacement	A	Annex A	01/02/2019	
					CT2750	Implantable cardiac electrical stimulators and associated devices	F1905	Device Revision or Replacement	A	Annex A	01/02/2019	
					CT2866	Implantable cardiac electrical stimulators	F1905	Device Revision or Replacement	A	Annex A	01/02/2019	
					CT2985	Active	F1905	Device Revision or Replacement	A	Annex A	01/02/2019	
					CT3001	Implantable metal	F1905	Device Revision or Replacement	A	Annex A	01/02/2019	
					CT3002	Implantable synthetic polymer	F1905	Device Revision or Replacement	A	Annex A	01/02/2019	
					CT3006	Implantable inorganic material	F1905	Device Revision or Replacement	A	Annex A	01/02/2019	
					CT314	Electrically powered	F1905	Device Revision or Replacement	A	Annex A	01/02/2019	
					CT321	Long-term surgical invasive	F1905	Device Revision or Replacement	A	Annex A	01/02/2019	
					CT449	Generators and associated devices	F1905	Device Revision or Replacement	A	Annex A	01/02/2019	
					CT981	Single use	F1905	Device Revision or Replacement	A	Annex A	01/02/2019	
					CT983	Surgical invasive	F1905	Device Revision or Replacement	A	Annex A	01/02/2019	
8455	2019	All other reportable incidents	35223	Endocardial/interventricular septal pacing lead	CT1684	Cardiac pacemakers and associated devices	F1905	Device Revision or Replacement	A070905	High Sensing Threshold		
8456	2019	All other reportable incidents	35223	Endocardial/interventricular septal pacing lead	CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A072202	Low Impedance		
8457	2020	All other reportable incidents	36241	Pacemaker/defibrillator lead	CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A051201	Device Dislodged or Dislocated	20/02/2019	20/02/2019
8458	2020	All other reportable incidents	47270	Cardiac resynchronization therapy implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A150206	Difficult to Insert	20/01/2020	20/01/2020
							F26	No Health Consequences or Impact	A150206	Difficult to Insert	20/01/2020	20/01/2020

8459	2019	All other reportable incidents	47265	Dual-chamber implantable pacemaker, rate-responsive	CT141	Electrodes/Leads and associated devices	F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	15/02/2019		
					CT145	Cardiovascular devices	F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	15/02/2019		
					CT1684	Cardiac pacemakers and associated devices	F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	15/02/2019		
					CT1685	Cardiac pacemakers	F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	15/02/2019		
					CT233	Surgical	F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	15/02/2019		
					CT2406	Implantable	F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	15/02/2019		
					CT2407	Active implantable	F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	15/02/2019		
					CT257	Implantable cardiac pacemakers	F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	15/02/2019		
					CT2727	Active implant devices/ancillaries	F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	15/02/2019		
					CT2750	Implantable cardiac electrical stimulators and associated devices	F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	15/02/2019		
					CT2866	Implantable cardiac electrical stimulators	F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	15/02/2019		
					CT2985	Active	F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	15/02/2019		
					CT3001	Implantable metal	F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	15/02/2019		
					CT3002	Implantable synthetic polymer	F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	15/02/2019		
					CT3006	Implantable inorganic material	F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	15/02/2019		
					CT314	Electrically powered	F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	15/02/2019		
					CT321	Long-term surgical invasive	F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	15/02/2019		
					CT449	Generators and associated devices	F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	15/02/2019		
					CT981	Single use	F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	15/02/2019		
					CT983	Surgical invasive	F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	15/02/2019		
8460	2019	All other reportable incidents	47263	Cardiac resynchronization therapy implantable pacemaker	CT1684	Cardiac pacemakers and associated devices	F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	04/03/2019		
					CT1685	Cardiac pacemakers	F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	04/03/2019		
					CT257	Implantable cardiac pacemakers	F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	04/03/2019		
					CT2866	Implantable cardiac electrical stimulators	F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	04/03/2019		
8461	2020	All other reportable incidents	35223	Endocardial/interventricular septal pacing lead	CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A090801	Signal Artifact/Noise	22/11/2019	22/11/2019	
							F1905	Device Revision or Replacement	A090801	Signal Artifact/Noise	22/11/2019	22/11/2019	
							F26	No Health Consequences or Impact	A090801	Signal Artifact/Noise	22/11/2019	22/11/2019	
8462	2020	All other reportable incidents	35852	Single-chamber implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A070909	Oversensing			
8463	2019	All other reportable incidents	17882	Physiologic-monitoring defibrillation system	CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A07	Electrical /Electronic Property Problem			
							F26	No Health Consequences or Impact	A07	Electrical /Electronic Property Problem			
8464	2019	All other reportable incidents	17882	Physiologic-monitoring defibrillation system	CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070803	Failure to Power Up			
							F27	Problem Identified During Non-Clinical Procedure	A070803	Failure to Power Up			
8465	2019	All other reportable incidents	17882	Physiologic-monitoring defibrillation system	CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A07	Electrical /Electronic Property Problem			
							F27	Problem Identified During Non-Clinical Procedure	A07	Electrical /Electronic Property Problem			
8466	2019	All other reportable incidents	17882	Physiologic-monitoring defibrillation system	CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070802	Intermittent Loss of Power			
									A070803	Failure to Power Up			
							F26	No Health Consequences or Impact	A070802	Intermittent Loss of Power			
									A070803	Failure to Power Up			
8467	2019	All other reportable incidents	17882	Physiologic-monitoring defibrillation system	CT1684	Cardiac pacemakers and associated devices	E2401	Insufficient Information	A070802	Intermittent Loss of Power			
									A070803	Failure to Power Up			
							E2403	No Clinical Signs, Symptoms or Conditions	A070802	Intermittent Loss of Power			
									A070803	Failure to Power Up			
							F26	No Health Consequences or Impact	A070802	Intermittent Loss of Power			
									A070803	Failure to Power Up			
8468	2019	All other reportable incidents	47265	Dual-chamber implantable pacemaker, rate-responsive	CT141	Electrodes/Leads and associated devices	F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery			
					CT145	Cardiovascular devices	F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery			
					CT1684	Cardiac pacemakers and associated devices	F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery			
					CT1685	Cardiac pacemakers	F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery			
					CT233	Surgical	F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery			
					CT2406	Implantable	F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery			
					CT2407	Active implantable	F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery			
					CT257	Implantable cardiac pacemakers	F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery			
					CT2727	Active implant devices/ancillaries	F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery			
					CT2750	Implantable cardiac electrical stimulators and associated devices	F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery			
					CT2866	Implantable cardiac electrical stimulators	F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery			
					CT2985	Active	F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery			
					CT3001	Implantable metal	F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery			
					CT3002	Implantable synthetic polymer	F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery			
					CT3006	Implantable inorganic material	F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery			
					CT314	Electrically powered	F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery			
					CT321	Long-term surgical invasive	F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery			
					CT449	Generators and associated devices	F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery			
					CT981	Single use	F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery			
					CT983	Surgical invasive	F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery			
8469	2019	All other reportable incidents	17882	Physiologic-monitoring defibrillation system	CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A071301	Failure to Deliver Shock/Stimulation			
							F21	Unexpected Deterioration	A071301	Failure to Deliver Shock/Stimulation			
							F26	No Health Consequences or Impact	A071301	Failure to Deliver Shock/Stimulation			
8470	2019	All other reportable incidents	17882	Physiologic-monitoring defibrillation system	CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A071301	Failure to Deliver Shock/Stimulation			
							F27	Problem Identified During Non-Clinical Procedure	A071301	Failure to Deliver Shock/Stimulation			
8471	2019	All other reportable incidents	47270	Cardiac resynchronization therapy implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A070908	Failure to Sense			
8472	2019	All other reportable incidents	17882	Physiologic-monitoring defibrillation system	CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070802	Intermittent Loss of Power			
									A0719	Unexpected Shutdown			
							F27	Problem Identified During Non-Clinical Procedure	A070802	Intermittent Loss of Power			
									A0719	Unexpected Shutdown			
8473	2019	All other reportable incidents	17882	Physiologic-monitoring defibrillation system	CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A07	Electrical /Electronic Property Problem			
							F27	Problem Identified During Non-Clinical Procedure	A07	Electrical /Electronic Property Problem			
8474	2019	All other reportable incidents	17882	Physiologic-monitoring defibrillation system	CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070908	Failure to Sense			
							F05	Delay to Treatment/ Therapy	A070908	Failure to Sense			
							F26	No Health Consequences or Impact	A070908	Failure to Sense			
8475	2019	All other reportable incidents	17882	Physiologic-monitoring defibrillation system	CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070903	Failure to Analyze Signal			
							F05	Delay to Treatment/ Therapy	A070903	Failure to Analyze Signal			
							F26	No Health Consequences or Impact	A070903	Failure to Analyze Signal			
8476	2019	All other reportable incidents	17882	Physiologic-monitoring defibrillation system	CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A07	Electrical /Electronic Property Problem			
							F27	Problem Identified During Non-Clinical Procedure	A07	Electrical /Electronic Property Problem			
8477	2019	All other reportable incidents	17882	Physiologic-monitoring defibrillation system	CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070908	Failure to Sense			

							F26	No Health Consequences or Impact	A070908	Failure to Sense		
8478	2019	All other reportable incidents	17882	Physiologic-monitoring defibrillation system	CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070802	Intermittent Loss of Power		
							F26	No Health Consequences or Impact	A070802	Intermittent Loss of Power		
8479	2019	Death	17882	Physiologic-monitoring defibrillation system	CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070908	Failure to Sense		
							F02	Death	A070908	Failure to Sense		
8480	2019	All other reportable incidents	17882	Physiologic-monitoring defibrillation system	CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070802	Intermittent Loss of Power		
							F26	No Health Consequences or Impact	A070802	Intermittent Loss of Power		
8481	2019	All other reportable incidents	17882	Physiologic-monitoring defibrillation system	CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A07	Electrical /Electronic Property Problem		
							F27	Problem Identified During Non-Clinical Procedure	A07	Electrical /Electronic Property Problem		
8482	2019	Death	17882	Physiologic-monitoring defibrillation system	CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070908	Failure to Sense		
							F02	Death	A070908	Failure to Sense		
							F1001	Absence of Treatment	A070908	Failure to Sense		
8483	2019	All other reportable incidents	17882	Physiologic-monitoring defibrillation system	CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A0719	Unexpected Shutdown		
							F26	No Health Consequences or Impact	A0719	Unexpected Shutdown		
8484	2019	All other reportable incidents	17882	Physiologic-monitoring defibrillation system	CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A071301	Failure to Deliver Shock/Stimulation		
							F05	Delay to Treatment/ Therapy	A071301	Failure to Deliver Shock/Stimulation		
							F26	No Health Consequences or Impact	A071301	Failure to Deliver Shock/Stimulation		
8485	2019	All other reportable incidents	17882	Physiologic-monitoring defibrillation system	CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A0719	Unexpected Shutdown		
							F27	Problem Identified During Non-Clinical Procedure	A0719	Unexpected Shutdown		
8486	2019	All other reportable incidents	37265	Dual-chamber implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A0709	Device Sensing Problem		
							F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A0709	Device Sensing Problem		
8487	2019	All other reportable incidents	36241	Pacemaker/defibrillator lead	CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A010402	Migration	11/02/2019	
8488	2020	All other reportable incidents	17882	Physiologic-monitoring defibrillation system	CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A071301	Failure to Deliver Shock/Stimulation		
8489	2020	All other reportable incidents	35852	Single-chamber implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	31/12/2019	31/12/2019
							F1901	Additional Surgery	A070504	Premature Discharge of Battery	31/12/2019	31/12/2019
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	31/12/2019	31/12/2019
8490	2020	All other reportable incidents	37265	Dual-chamber implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	05/02/2021	05/02/2021
									A0712	Pacing Problem	05/02/2021	05/02/2021
									A071302	Inappropriate/Inadequate Shock/Stimulation	05/02/2021	05/02/2021
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	05/02/2021	05/02/2021
									A0712	Pacing Problem	05/02/2021	05/02/2021
									A071302	Inappropriate/Inadequate Shock/Stimulation	05/02/2021	05/02/2021
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	05/02/2021	05/02/2021
									A0712	Pacing Problem	05/02/2021	05/02/2021
									A071302	Inappropriate/Inadequate Shock/Stimulation	05/02/2021	05/02/2021
8491	2020	All other reportable incidents	47267	Single-chamber implantable pacemaker, rate-responsive	CT141	Electrodes/Leads and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	24/09/2019	24/09/2019
					CT145	Cardiovascular devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	24/09/2019	24/09/2019
					CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	24/09/2019	24/09/2019
					CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	24/09/2019	24/09/2019
					CT233	Surgical	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	24/09/2019	24/09/2019
					CT2406	Implantable	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	24/09/2019	24/09/2019
					CT2407	Active implantable	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	24/09/2019	24/09/2019
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	24/09/2019	24/09/2019
					CT2727	Active implant devices/ancillaries	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	24/09/2019	24/09/2019
					CT2750	Implantable cardiac electrical stimulators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	24/09/2019	24/09/2019
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	24/09/2019	24/09/2019
					CT2985	Active	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	24/09/2019	24/09/2019
					CT3001	Implantable metal	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	24/09/2019	24/09/2019
					CT3002	Implantable synthetic polymer	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	24/09/2019	24/09/2019
					CT3006	Implantable inorganic material	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	24/09/2019	24/09/2019
					CT314	Electrically powered	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	24/09/2019	24/09/2019
					CT321	Long-term surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	24/09/2019	24/09/2019
					CT449	Generators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	24/09/2019	24/09/2019
					CT981	Single use	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	24/09/2019	24/09/2019
					CT983	Surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	24/09/2019	24/09/2019
8492	2020	All other reportable incidents	36241	Pacemaker/defibrillator lead	CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A072201	High Impedance		
8493	2020	All other reportable incidents	35852	Single-chamber implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A072201	High Impedance		
8494	2020	All other reportable incidents	36241	Pacemaker/defibrillator lead	CT1684	Cardiac pacemakers and associated devices	E0112	Dizziness	A051201	Device Dislodged or Dislocated	20/12/2019	20/12/2019
							E0119	Loss of Consciousness	A051201	Device Dislodged or Dislocated	20/12/2019	20/12/2019
							E2403	No Clinical Signs, Symptoms or Conditions	A051201	Device Dislodged or Dislocated	20/12/2019	20/12/2019
							F19	Surgical Intervention	A051201	Device Dislodged or Dislocated	20/12/2019	20/12/2019
8495	2019	All other reportable incidents	37265	Dual-chamber implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	06/03/2019	
8496	2019	All other reportable incidents	47267	Single-chamber implantable pacemaker, rate-responsive	CT141	Electrodes/Leads and associated devices	F1905	Device Revision or Replacement	A071205	Pacemaker Found in Back-Up Mode		
					CT145	Cardiovascular devices	F1905	Device Revision or Replacement	A071205	Pacemaker Found in Back-Up Mode		
					CT1684	Cardiac pacemakers and associated devices	F1905	Device Revision or Replacement	A071205	Pacemaker Found in Back-Up Mode		
					CT1685	Cardiac pacemakers	F1905	Device Revision or Replacement	A071205	Pacemaker Found in Back-Up Mode		
					CT233	Surgical	F1905	Device Revision or Replacement	A071205	Pacemaker Found in Back-Up Mode		
					CT2406	Implantable	F1905	Device Revision or Replacement	A071205	Pacemaker Found in Back-Up Mode		
					CT2407	Active implantable	F1905	Device Revision or Replacement	A071205	Pacemaker Found in Back-Up Mode		
					CT257	Implantable cardiac pacemakers	F1905	Device Revision or Replacement	A071205	Pacemaker Found in Back-Up Mode		
					CT2727	Active implant devices/ancillaries	F1905	Device Revision or Replacement	A071205	Pacemaker Found in Back-Up Mode		
					CT2750	Implantable cardiac electrical stimulators and associated devices	F1905	Device Revision or Replacement	A071205	Pacemaker Found in Back-Up Mode		
					CT2866	Implantable cardiac electrical stimulators	F1905	Device Revision or Replacement	A071205	Pacemaker Found in Back-Up Mode		
					CT2985	Active	F1905	Device Revision or Replacement	A071205	Pacemaker Found in Back-Up Mode		
					CT3001	Implantable metal	F1905	Device Revision or Replacement	A071205	Pacemaker Found in Back-Up Mode		
					CT3002	Implantable synthetic polymer	F1905	Device Revision or Replacement	A071205	Pacemaker Found in Back-Up Mode		
					CT3006	Implantable inorganic material	F1905	Device Revision or Replacement	A071205	Pacemaker Found in Back-Up Mode		
					CT314	Electrically powered	F1905	Device Revision or Replacement	A071205	Pacemaker Found in Back-Up Mode		
					CT321	Long-term surgical invasive	F1905	Device Revision or Replacement	A071205	Pacemaker Found in Back-Up Mode		
					CT449	Generators and associated devices	F1905	Device Revision or Replacement	A071205	Pacemaker Found in Back-Up Mode		
					CT981	Single use	F1905	Device Revision or Replacement	A071205	Pacemaker Found in Back-Up Mode		
					CT983	Surgical invasive	F1905	Device Revision or Replacement	A071205	Pacemaker Found in Back-Up Mode		
8497	2019	All other reportable incidents	35852	Single-chamber implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	05/03/2019	
8498	2019	All other reportable incidents	35852	Single-chamber implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A070909	Oversensing		
8499	2019	All other reportable incidents	36241	Pacemaker/defibrillator lead	CT1684	Cardiac pacemakers and associated devices	E2109	Implant Pain	A0104	Migration or Expulsion of Device	06/02/2019	
							F1905	Device Revision or Replacement	A0104	Migration or Expulsion of Device	06/02/2019	

8500	2020	All other reportable incidents	37265	Dual-chamber implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	06/01/2020	06/01/2020
							F1901	Additional Surgery	A070504	Premature Discharge of Battery	06/01/2020	06/01/2020
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	06/01/2020	06/01/2020
8501	2019	All other reportable incidents	47265	Dual-chamber implantable pacemaker, rate-responsive	CT141	Electrodes/Leads and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A071206	Pacing Asynchronously		
					CT145	Cardiovascular devices	E2403	No Clinical Signs, Symptoms or Conditions	A071206	Pacing Asynchronously		
					CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A071206	Pacing Asynchronously		
					CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A071206	Pacing Asynchronously		
					CT233	Surgical	E2403	No Clinical Signs, Symptoms or Conditions	A071206	Pacing Asynchronously		
					CT2406	Implantable	E2403	No Clinical Signs, Symptoms or Conditions	A071206	Pacing Asynchronously		
					CT2407	Active implantable	E2403	No Clinical Signs, Symptoms or Conditions	A071206	Pacing Asynchronously		
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A071206	Pacing Asynchronously		
					CT2727	Active implant devices/ancillaries	E2403	No Clinical Signs, Symptoms or Conditions	A071206	Pacing Asynchronously		
					CT2750	Implantable cardiac electrical stimulators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A071206	Pacing Asynchronously		
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A071206	Pacing Asynchronously		
					CT2985	Active	E2403	No Clinical Signs, Symptoms or Conditions	A071206	Pacing Asynchronously		
					CT3001	Implantable metal	E2403	No Clinical Signs, Symptoms or Conditions	A071206	Pacing Asynchronously		
					CT3002	Implantable synthetic polymer	E2403	No Clinical Signs, Symptoms or Conditions	A071206	Pacing Asynchronously		
					CT3006	Implantable inorganic material	E2403	No Clinical Signs, Symptoms or Conditions	A071206	Pacing Asynchronously		
					CT314	Electrically powered	E2403	No Clinical Signs, Symptoms or Conditions	A071206	Pacing Asynchronously		
					CT321	Long-term surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A071206	Pacing Asynchronously		
					CT449	Generators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A071206	Pacing Asynchronously		
					CT981	Single use	E2403	No Clinical Signs, Symptoms or Conditions	A071206	Pacing Asynchronously		
					CT983	Surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A071206	Pacing Asynchronously		
8502	2019	All other reportable incidents	36241	Pacemaker/defibrillator lead	CT1684	Cardiac pacemakers and associated devices	F1905	Device Revision or Replacement	A0510	Retraction Problem	15/02/2019	
8503	2019	All other reportable incidents	37265	Dual-chamber implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A070909	Oversensing		
8504	2019	All other reportable incidents	47265	Dual-chamber implantable pacemaker, rate-responsive	CT141	Electrodes/Leads and associated devices	F1905	Device Revision or Replacement	A0705	Battery Problem		
					CT145	Cardiovascular devices	F1905	Device Revision or Replacement	A0705	Battery Problem		
					CT1684	Cardiac pacemakers and associated devices	F1905	Device Revision or Replacement	A0705	Battery Problem		
					CT1685	Cardiac pacemakers	F1905	Device Revision or Replacement	A0705	Battery Problem		
					CT233	Surgical	F1905	Device Revision or Replacement	A0705	Battery Problem		
					CT2406	Implantable	F1905	Device Revision or Replacement	A0705	Battery Problem		
					CT2407	Active implantable	F1905	Device Revision or Replacement	A0705	Battery Problem		
					CT257	Implantable cardiac pacemakers	F1905	Device Revision or Replacement	A0705	Battery Problem		
					CT2727	Active implant devices/ancillaries	F1905	Device Revision or Replacement	A0705	Battery Problem		
					CT2750	Implantable cardiac electrical stimulators and associated devices	F1905	Device Revision or Replacement	A0705	Battery Problem		
					CT2866	Implantable cardiac electrical stimulators	F1905	Device Revision or Replacement	A0705	Battery Problem		
					CT2985	Active	F1905	Device Revision or Replacement	A0705	Battery Problem		
					CT3001	Implantable metal	F1905	Device Revision or Replacement	A0705	Battery Problem		
					CT3002	Implantable synthetic polymer	F1905	Device Revision or Replacement	A0705	Battery Problem		
					CT3006	Implantable inorganic material	F1905	Device Revision or Replacement	A0705	Battery Problem		
					CT314	Electrically powered	F1905	Device Revision or Replacement	A0705	Battery Problem		
					CT321	Long-term surgical invasive	F1905	Device Revision or Replacement	A0705	Battery Problem		
					CT449	Generators and associated devices	F1905	Device Revision or Replacement	A0705	Battery Problem		
					CT981	Single use	F1905	Device Revision or Replacement	A0705	Battery Problem		
					CT983	Surgical invasive	F1905	Device Revision or Replacement	A0705	Battery Problem		
8505	2019	All other reportable incidents	36241	Pacemaker/defibrillator lead	CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070909	Oversensing		
8506	2019	All other reportable incidents	47265	Dual-chamber implantable pacemaker, rate-responsive	CT141	Electrodes/Leads and associated devices	F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A070909	Oversensing		
					CT145	Cardiovascular devices	F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A070909	Oversensing		
					CT1684	Cardiac pacemakers and associated devices	F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A070909	Oversensing		
					CT1685	Cardiac pacemakers	F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A070909	Oversensing		
					CT233	Surgical	F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A070909	Oversensing		
					CT2406	Implantable	F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A070909	Oversensing		
					CT2407	Active implantable	F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A070909	Oversensing		
					CT257	Implantable cardiac pacemakers	F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A070909	Oversensing		
					CT2727	Active implant devices/ancillaries	F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A070909	Oversensing		
					CT2750	Implantable cardiac electrical stimulators and associated devices	F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A070909	Oversensing		
					CT2866	Implantable cardiac electrical stimulators	F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A070909	Oversensing		
					CT2985	Active	F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A070909	Oversensing		
					CT3001	Implantable metal	F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A070909	Oversensing		
					CT3002	Implantable synthetic polymer	F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A070909	Oversensing		
					CT3006	Implantable inorganic material	F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A070909	Oversensing		
					CT314	Electrically powered	F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A070909	Oversensing		
					CT321	Long-term surgical invasive	F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A070909	Oversensing		
					CT449	Generators and associated devices	F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A070909	Oversensing		
					CT981	Single use	F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A070909	Oversensing		
					CT983	Surgical invasive	F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A070909	Oversensing		
8507	2019	All other reportable incidents	36241	Pacemaker/defibrillator lead	CT1684	Cardiac pacemakers and associated devices	F1905	Device Revision or Replacement	A010402	Migration		
8508	2019	All other reportable incidents	36241	Pacemaker/defibrillator lead	CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A0510	Retraction Problem	13/02/2019	
8509	2020	All other reportable incidents	35224	External pacemaker, epicardial pacing	CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070801	Complete Loss of Power		
					CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A070801	Complete Loss of Power		
8510	2019	All other reportable incidents	47270	Cardiac resynchronization therapy implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	F1905	Device Revision or Replacement	A110401	Operating System Becomes Nonfunctional	06/02/2019	
8511	2020	All other reportable incidents	35852	Single-chamber implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A070909	Oversensing		
									A071302	Inappropriate/Inadequate Shock/Stimulation		
8512	2019	All other reportable incidents	47270	Cardiac resynchronization therapy implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	27/02/2019	
8513	2020	All other reportable incidents	37265	Dual-chamber implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A071302	Inappropriate/Inadequate Shock/Stimulation		
8514	2019	All other reportable incidents	35852	Single-chamber implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	28/02/2019	
8515	2020	All other reportable incidents	36241	Pacemaker/defibrillator lead	CT1684	Cardiac pacemakers and associated devices	E0112	Dizziness	A051201	Device Dislodged or Dislocated	20/12/2019	20/12/2019
									A070101	Failure to Capture	20/12/2019	20/12/2019
									A0712	Pacing Problem	20/12/2019	20/12/2019
							F1001	Absence of Treatment	A051201	Device Dislodged or Dislocated	20/12/2019	20/12/2019
									A070101	Failure to Capture	20/12/2019	20/12/2019
									A0712	Pacing Problem	20/12/2019	20/12/2019
8516	2019	All other reportable incidents	36241	Pacemaker/defibrillator lead	CT1684	Cardiac pacemakers and associated devices	E2114	Perforation	A010402	Migration		
							F1905	Device Revision or Replacement	A010402	Migration		

8517	2019	All other reportable incidents	36241	Pacemaker/defibrillator lead	CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A0510	Retraction Problem	13/02/2019		
8518	2019	All other reportable incidents	47265	Dual-chamber implantable pacemaker, rate-responsive	CT141	Electrodes/Leads and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070910	Undersensing			
					CT145	Cardiovascular devices	E2403	No Clinical Signs, Symptoms or Conditions	A070910	Undersensing			
					CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070910	Undersensing			
					CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A070910	Undersensing			
					CT233	Surgical	E2403	No Clinical Signs, Symptoms or Conditions	A070910	Undersensing			
					CT2406	Implantable	E2403	No Clinical Signs, Symptoms or Conditions	A070910	Undersensing			
					CT2407	Active implantable	E2403	No Clinical Signs, Symptoms or Conditions	A070910	Undersensing			
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A070910	Undersensing			
					CT2727	Active implant devices/ancillaries	E2403	No Clinical Signs, Symptoms or Conditions	A070910	Undersensing			
					CT2750	Implantable cardiac electrical stimulators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070910	Undersensing			
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A070910	Undersensing			
					CT2985	Active	E2403	No Clinical Signs, Symptoms or Conditions	A070910	Undersensing			
					CT3001	Implantable metal	E2403	No Clinical Signs, Symptoms or Conditions	A070910	Undersensing			
					CT3002	Implantable synthetic polymer	E2403	No Clinical Signs, Symptoms or Conditions	A070910	Undersensing			
					CT3006	Implantable inorganic material	E2403	No Clinical Signs, Symptoms or Conditions	A070910	Undersensing			
					CT314	Electrically powered	E2403	No Clinical Signs, Symptoms or Conditions	A070910	Undersensing			
					CT321	Long-term surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A070910	Undersensing			
					CT449	Generators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070910	Undersensing			
					CT981	Single use	E2403	No Clinical Signs, Symptoms or Conditions	A070910	Undersensing			
					CT983	Surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A070910	Undersensing			
8519	2019	All other reportable incidents	36241	Pacemaker/defibrillator lead	CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A0510	Retraction Problem			
8520	2019	All other reportable incidents	36241	Pacemaker/defibrillator lead	CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A0510	Retraction Problem	13/02/2019		
8521	2019	All other reportable incidents	47265	Dual-chamber implantable pacemaker, rate-responsive	CT141	Electrodes/Leads and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery			
					CT145	Cardiovascular devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery			
					CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery			
					CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery			
					CT233	Surgical	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery			
					CT2406	Implantable	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery			
					CT2407	Active implantable	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery			
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery			
					CT2727	Active implant devices/ancillaries	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery			
					CT2750	Implantable cardiac electrical stimulators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery			
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery			
					CT2985	Active	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery			
					CT3001	Implantable metal	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery			
					CT3002	Implantable synthetic polymer	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery			
					CT3006	Implantable inorganic material	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery			
					CT314	Electrically powered	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery			
					CT321	Long-term surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery			
					CT449	Generators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery			
					CT981	Single use	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery			
					CT983	Surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery			
8522	2019	All other reportable incidents	17882	Physiologic-monitoring defibrillation system	CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A0203	Defective Device			
8523	2019	All other reportable incidents	17882	Physiologic-monitoring defibrillation system	CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070802	Intermittent Loss of Power			
8524	2019	All other reportable incidents	47265	Dual-chamber implantable pacemaker, rate-responsive	CT141	Electrodes/Leads and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery			
					CT145	Cardiovascular devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery			
					CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery			
					CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery			
					CT233	Surgical	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery			
					CT2406	Implantable	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery			
					CT2407	Active implantable	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery			
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery			
					CT2727	Active implant devices/ancillaries	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery			
					CT2750	Implantable cardiac electrical stimulators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery			
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery			
					CT2985	Active	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery			
					CT3001	Implantable metal	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery			
					CT3002	Implantable synthetic polymer	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery			
					CT3006	Implantable inorganic material	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery			
					CT314	Electrically powered	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery			
					CT321	Long-term surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery			
					CT449	Generators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery			
					CT981	Single use	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery			
					CT983	Surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery			
8525	2019	All other reportable incidents	17882	Physiologic-monitoring defibrillation system	CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A1205	Loose or Intermittent Connection			
8526	2019	All other reportable incidents	36241	Pacemaker/defibrillator lead	CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070909	Oversensing			
8527	2019	All other reportable incidents	37265	Dual-chamber implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A070909	Oversensing			
8528	2020	Death	47270	Cardiac resynchronization therapy implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E0601	Arrhythmia	A0701	Capturing Problem			
									A070909	Oversensing			
									A070910	Undersensing			
									A071301	Failure to Deliver Shock/Stimulation			
							F02	Death	A0701	Capturing Problem			
									A070909	Oversensing			
									A070910	Undersensing			
									A071301	Failure to Deliver Shock/Stimulation			
							F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A0701	Capturing Problem			
									A070909	Oversensing			
									A070910	Undersensing			
									A071301	Failure to Deliver Shock/Stimulation			
8529	2020	All other reportable incidents	47270	Cardiac resynchronization therapy implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	06/01/2020	06/01/2020	
							F1901	Additional Surgery	A070504	Premature Discharge of Battery	06/01/2020	06/01/2020	
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	06/01/2020	06/01/2020	
8530	2020	All other reportable incidents	35852	Single-chamber implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	10/01/2020	10/01/2020	
							F1901	Additional Surgery	A070504	Premature Discharge of Battery	10/01/2020	10/01/2020	

							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	10/01/2020	10/01/2020
8531	2020	All other reportable incidents	36241	Pacemaker/defibrillator lead	CT1684	Cardiac pacemakers and associated devices	F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A070909	Oversensing	08/01/2020	08/01/2020
									A071302	Inappropriate/Inadequate Shock/Stimulation	08/01/2020	08/01/2020
8532	2019	All other reportable incidents	17882	Physiologic-monitoring defibrillation system	CT1684	Cardiac pacemakers and associated devices	F05	Delay to Treatment/ Therapy	A070908	Failure to Sense		
8533	2019	All other reportable incidents	35223	Endocardial/interventricular septal pacing lead	CT1684	Cardiac pacemakers and associated devices	E2007	Fall	A0510	Retraction Problem		
							F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A0510	Retraction Problem		
							F1905	Device Revision or Replacement	A0510	Retraction Problem		
8534	2019	All other reportable incidents	35223	Endocardial/interventricular septal pacing lead	CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A072202	Low Impedance		
8535	2019	All other reportable incidents	47265	Dual-chamber implantable pacemaker, rate-responsive	CT141	Electrodes/Leads and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A160602	Premature End-of-Life Indicator		
					CT145	Cardiovascular devices	E2403	No Clinical Signs, Symptoms or Conditions	A160602	Premature End-of-Life Indicator		
					CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A160602	Premature End-of-Life Indicator		
					CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A160602	Premature End-of-Life Indicator		
					CT233	Surgical	E2403	No Clinical Signs, Symptoms or Conditions	A160602	Premature End-of-Life Indicator		
					CT2406	Implantable	E2403	No Clinical Signs, Symptoms or Conditions	A160602	Premature End-of-Life Indicator		
					CT2407	Active implantable	E2403	No Clinical Signs, Symptoms or Conditions	A160602	Premature End-of-Life Indicator		
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A160602	Premature End-of-Life Indicator		
					CT2727	Active implant devices/ancillaries	E2403	No Clinical Signs, Symptoms or Conditions	A160602	Premature End-of-Life Indicator		
					CT2750	Implantable cardiac electrical stimulators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A160602	Premature End-of-Life Indicator		
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A160602	Premature End-of-Life Indicator		
					CT2985	Active	E2403	No Clinical Signs, Symptoms or Conditions	A160602	Premature End-of-Life Indicator		
					CT3001	Implantable metal	E2403	No Clinical Signs, Symptoms or Conditions	A160602	Premature End-of-Life Indicator		
					CT3002	Implantable synthetic polymer	E2403	No Clinical Signs, Symptoms or Conditions	A160602	Premature End-of-Life Indicator		
					CT3006	Implantable inorganic material	E2403	No Clinical Signs, Symptoms or Conditions	A160602	Premature End-of-Life Indicator		
					CT314	Electrically powered	E2403	No Clinical Signs, Symptoms or Conditions	A160602	Premature End-of-Life Indicator		
					CT321	Long-term surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A160602	Premature End-of-Life Indicator		
					CT449	Generators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A160602	Premature End-of-Life Indicator		
					CT981	Single use	E2403	No Clinical Signs, Symptoms or Conditions	A160602	Premature End-of-Life Indicator		
					CT983	Surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A160602	Premature End-of-Life Indicator		
8536	2020	All other reportable incidents	37265	Dual-chamber implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	21/01/2020	21/01/2020
							F1901	Additional Surgery	A070504	Premature Discharge of Battery	21/01/2020	21/01/2020
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	21/01/2020	21/01/2020
8537	2019	All other reportable incidents	47270	Cardiac resynchronization therapy implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	20/03/2019	20/03/2019
8538	2020	All other reportable incidents	47265	Dual-chamber implantable pacemaker, rate-responsive	CT141	Electrodes/Leads and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
					CT145	Cardiovascular devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
					CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
					CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
					CT233	Surgical	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
					CT2406	Implantable	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
					CT2407	Active implantable	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
					CT2727	Active implant devices/ancillaries	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
					CT2750	Implantable cardiac electrical stimulators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
					CT2985	Active	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
					CT3001	Implantable metal	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
					CT3002	Implantable synthetic polymer	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
					CT3006	Implantable inorganic material	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
					CT314	Electrically powered	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
					CT321	Long-term surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
					CT449	Generators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
					CT981	Single use	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
					CT983	Surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
8539	2020	All other reportable incidents	35852	Single-chamber implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A071302	Inappropriate/Inadequate Shock/Stimulation		
8540	2020	All other reportable incidents	47263	Cardiac resynchronization therapy implantable pacemaker	CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F19	Surgical Intervention	A070504	Premature Discharge of Battery		
					CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F19	Surgical Intervention	A070504	Premature Discharge of Battery		
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F19	Surgical Intervention	A070504	Premature Discharge of Battery		
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F19	Surgical Intervention	A070504	Premature Discharge of Battery		
8541	2020	All other reportable incidents	37265	Dual-chamber implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	16/01/2020	16/01/2020
									A090201	Device Displays Incorrect Message	16/01/2020	16/01/2020
8542	2019	All other reportable incidents	36241	Pacemaker/defibrillator lead	CT1684	Cardiac pacemakers and associated devices	F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A071302	Inappropriate/Inadequate Shock/Stimulation		
8543	2019	All other reportable incidents	35852	Single-chamber implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A071302	Inappropriate/Inadequate Shock/Stimulation		
8544	2019	All other reportable incidents	36241	Pacemaker/defibrillator lead	CT1684	Cardiac pacemakers and associated devices	F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A	Annex A	22/02/2019	
							F1905	Device Revision or Replacement	A	Annex A	22/02/2019	
8545	2019	All other reportable incidents	35852	Single-chamber implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A071302	Inappropriate/Inadequate Shock/Stimulation	22/02/2019	
							F1905	Device Revision or Replacement	A071302	Inappropriate/Inadequate Shock/Stimulation	22/02/2019	
8546	2019	All other reportable incidents	47267	Single-chamber implantable pacemaker, rate-responsive	CT141	Electrodes/Leads and associated devices	F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery		
					CT145	Cardiovascular devices	F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery		
					CT1684	Cardiac pacemakers and associated devices	F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery		
					CT1685	Cardiac pacemakers	F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery		
					CT233	Surgical	F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery		
					CT2406	Implantable	F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery		
					CT2407	Active implantable	F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery		
					CT257	Implantable cardiac pacemakers	F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery		
					CT2727	Active implant devices/ancillaries	F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery		
					CT2750	Implantable cardiac electrical stimulators and associated devices	F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery		
					CT2866	Implantable cardiac electrical stimulators	F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery		
					CT2985	Active	F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery		
					CT3001	Implantable metal	F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery		
					CT3002	Implantable synthetic polymer	F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery		
					CT3006	Implantable inorganic material	F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery		

					CT314	Electrically powered	F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery		
					CT321	Long-term surgical invasive	F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery		
					CT449	Generators and associated devices	F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery		
					CT981	Single use	F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery		
					CT983	Surgical invasive	F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery		
8547	2019	All other reportable incidents	47265	Dual-chamber implantable pacemaker, rate-responsive	CT141	Electrodes/Leads and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
					CT145	Cardiovascular devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
					CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
					CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
					CT233	Surgical	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
					CT2406	Implantable	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
					CT2407	Active implantable	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
					CT2727	Active implant devices/ancillaries	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
					CT2750	Implantable cardiac electrical stimulators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
					CT2985	Active	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
					CT3001	Implantable metal	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
					CT3002	Implantable synthetic polymer	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
					CT3006	Implantable inorganic material	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
					CT314	Electrically powered	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
					CT321	Long-term surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
					CT449	Generators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
					CT981	Single use	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
					CT983	Surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
8548	2020	All other reportable incidents	35223	Endocardial/interventricular septal pacing lead	CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070101	Failure to Capture		
									A070909	Oversensing		
							F1901	Additional Surgery	A070101	Failure to Capture		
									A070909	Oversensing		
							F1905	Device Revision or Replacement	A070101	Failure to Capture		
									A070909	Oversensing		
8549	2020	All other reportable incidents	35223	Endocardial/interventricular septal pacing lead	CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070908	Failure to Sense		
									A070910	Undersensing		
							F1901	Additional Surgery	A070908	Failure to Sense		
									A070910	Undersensing		
							F1905	Device Revision or Replacement	A070908	Failure to Sense		
									A070910	Undersensing		
8550	2020	All other reportable incidents	35852	Single-chamber implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
8551	2020	All other reportable incidents	47267	Single-chamber implantable pacemaker, rate-responsive	CT141	Electrodes/Leads and associated devices	F1905	Device Revision or Replacement	A24	No Apparent Device Problem	30/10/2019	30/10/2019
					CT145	Cardiovascular devices	F1905	Device Revision or Replacement	A24	No Apparent Device Problem	30/10/2019	30/10/2019
					CT1684	Cardiac pacemakers and associated devices	F1905	Device Revision or Replacement	A24	No Apparent Device Problem	30/10/2019	30/10/2019
					CT1685	Cardiac pacemakers	F1905	Device Revision or Replacement	A24	No Apparent Device Problem	30/10/2019	30/10/2019
					CT233	Surgical	F1905	Device Revision or Replacement	A24	No Apparent Device Problem	30/10/2019	30/10/2019
					CT2406	Implantable	F1905	Device Revision or Replacement	A24	No Apparent Device Problem	30/10/2019	30/10/2019
					CT2407	Active implantable	F1905	Device Revision or Replacement	A24	No Apparent Device Problem	30/10/2019	30/10/2019
					CT257	Implantable cardiac pacemakers	F1905	Device Revision or Replacement	A24	No Apparent Device Problem	30/10/2019	30/10/2019
					CT2727	Active implant devices/ancillaries	F1905	Device Revision or Replacement	A24	No Apparent Device Problem	30/10/2019	30/10/2019
					CT2750	Implantable cardiac electrical stimulators and associated devices	F1905	Device Revision or Replacement	A24	No Apparent Device Problem	30/10/2019	30/10/2019
					CT2866	Implantable cardiac electrical stimulators	F1905	Device Revision or Replacement	A24	No Apparent Device Problem	30/10/2019	30/10/2019
					CT2985	Active	F1905	Device Revision or Replacement	A24	No Apparent Device Problem	30/10/2019	30/10/2019
					CT3001	Implantable metal	F1905	Device Revision or Replacement	A24	No Apparent Device Problem	30/10/2019	30/10/2019
					CT3002	Implantable synthetic polymer	F1905	Device Revision or Replacement	A24	No Apparent Device Problem	30/10/2019	30/10/2019
					CT3006	Implantable inorganic material	F1905	Device Revision or Replacement	A24	No Apparent Device Problem	30/10/2019	30/10/2019
					CT314	Electrically powered	F1905	Device Revision or Replacement	A24	No Apparent Device Problem	30/10/2019	30/10/2019
					CT321	Long-term surgical invasive	F1905	Device Revision or Replacement	A24	No Apparent Device Problem	30/10/2019	30/10/2019
					CT449	Generators and associated devices	F1905	Device Revision or Replacement	A24	No Apparent Device Problem	30/10/2019	30/10/2019
					CT981	Single use	F1905	Device Revision or Replacement	A24	No Apparent Device Problem	30/10/2019	30/10/2019
					CT983	Surgical invasive	F1905	Device Revision or Replacement	A24	No Apparent Device Problem	30/10/2019	30/10/2019
8552	2020	All other reportable incidents	47265	Dual-chamber implantable pacemaker, rate-responsive	CT141	Electrodes/Leads and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	01/03/2020	01/03/2020
					CT145	Cardiovascular devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	01/03/2020	01/03/2020
					CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	01/03/2020	01/03/2020
					CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	01/03/2020	01/03/2020
					CT233	Surgical	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	01/03/2020	01/03/2020
					CT2406	Implantable	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	01/03/2020	01/03/2020
					CT2407	Active implantable	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	01/03/2020	01/03/2020
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	01/03/2020	01/03/2020
					CT2727	Active implant devices/ancillaries	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	01/03/2020	01/03/2020
					CT2750	Implantable cardiac electrical stimulators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	01/03/2020	01/03/2020
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	01/03/2020	01/03/2020
					CT2985	Active	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	01/03/2020	01/03/2020
					CT3001	Implantable metal	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	01/03/2020	01/03/2020
					CT3002	Implantable synthetic polymer	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	01/03/2020	01/03/2020
					CT3006	Implantable inorganic material	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	01/03/2020	01/03/2020
					CT314	Electrically powered	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	01/03/2020	01/03/2020
					CT321	Long-term surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	01/03/2020	01/03/2020
					CT449	Generators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	01/03/2020	01/03/2020
					CT981	Single use	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	01/03/2020	01/03/2020
					CT983	Surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	01/03/2020	01/03/2020
8553	2020	All other reportable incidents	35852	Single-chamber implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	23/12/2019	23/12/2019
8554	2019	All other reportable incidents	35223	Endocardial/interventricular septal pacing lead	CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070102	High Capture Threshold		
									A072202	Low Impedance		
8555	2019	All other reportable incidents	35223	Endocardial/interventricular septal pacing lead	CT1684	Cardiac pacemakers and associated devices	F1905	Device Revision or Replacement	A0401	Break		
									A070909	Oversensing		
									A072202	Low Impedance		
8556	2019	All other reportable incidents	35224	External pacemaker, epicardial pacing	CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A0719	Unexpected Shutdown		

					CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A0719	Unexpected Shutdown			
8557	2019	All other reportable incidents	35852	Single-chamber implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A0712	Pacing Problem	06/12/2019	06/12/2019	
									A09	Output Problem	06/12/2019	06/12/2019	
8558	2019	All other reportable incidents	37265	Dual-chamber implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem			
									A070602	Delayed Charge Time			
8559	2020	All other reportable incidents	35223	Endocardial/interventricular septal pacing lead	CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070102	High Capture Threshold			
									A070909	Oversensing			
									A072201	High Impedance			
8560	2019	All other reportable incidents	35223	Endocardial/interventricular septal pacing lead	CT1684	Cardiac pacemakers and associated devices	F1905	Device Revision or Replacement	A070909	Oversensing	08/01/2019		
8561	2019	Death	47265	Dual-chamber implantable pacemaker, rate-responsive	CT141	Electrodes/Leads and associated devices	F21	Unexpected Deterioration	A110206	Application Program Version or Upgrade Problem	13/03/2019		
					CT145	Cardiovascular devices	F21	Unexpected Deterioration	A110206	Application Program Version or Upgrade Problem	13/03/2019		
					CT1684	Cardiac pacemakers and associated devices	F21	Unexpected Deterioration	A110206	Application Program Version or Upgrade Problem	13/03/2019		
					CT1685	Cardiac pacemakers	F21	Unexpected Deterioration	A110206	Application Program Version or Upgrade Problem	13/03/2019		
					CT233	Surgical	F21	Unexpected Deterioration	A110206	Application Program Version or Upgrade Problem	13/03/2019		
					CT2406	Implantable	F21	Unexpected Deterioration	A110206	Application Program Version or Upgrade Problem	13/03/2019		
					CT2407	Active implantable	F21	Unexpected Deterioration	A110206	Application Program Version or Upgrade Problem	13/03/2019		
					CT257	Implantable cardiac pacemakers	F21	Unexpected Deterioration	A110206	Application Program Version or Upgrade Problem	13/03/2019		
					CT2727	Active implant devices/ancillaries	F21	Unexpected Deterioration	A110206	Application Program Version or Upgrade Problem	13/03/2019		
					CT2750	Implantable cardiac electrical stimulators and associated devices	F21	Unexpected Deterioration	A110206	Application Program Version or Upgrade Problem	13/03/2019		
					CT2866	Implantable cardiac electrical stimulators	F21	Unexpected Deterioration	A110206	Application Program Version or Upgrade Problem	13/03/2019		
					CT2985	Active	F21	Unexpected Deterioration	A110206	Application Program Version or Upgrade Problem	13/03/2019		
					CT3001	Implantable metal	F21	Unexpected Deterioration	A110206	Application Program Version or Upgrade Problem	13/03/2019		
					CT3002	Implantable synthetic polymer	F21	Unexpected Deterioration	A110206	Application Program Version or Upgrade Problem	13/03/2019		
					CT3006	Implantable inorganic material	F21	Unexpected Deterioration	A110206	Application Program Version or Upgrade Problem	13/03/2019		
					CT314	Electrically powered	F21	Unexpected Deterioration	A110206	Application Program Version or Upgrade Problem	13/03/2019		
					CT321	Long-term surgical invasive	F21	Unexpected Deterioration	A110206	Application Program Version or Upgrade Problem	13/03/2019		
					CT449	Generators and associated devices	F21	Unexpected Deterioration	A110206	Application Program Version or Upgrade Problem	13/03/2019		
					CT981	Single use	F21	Unexpected Deterioration	A110206	Application Program Version or Upgrade Problem	13/03/2019		
					CT983	Surgical invasive	F21	Unexpected Deterioration	A110206	Application Program Version or Upgrade Problem	13/03/2019		
8562	2019	All other reportable incidents	35223	Endocardial/interventricular septal pacing lead	CT1684	Cardiac pacemakers and associated devices	F1905	Device Revision or Replacement	A070101	Failure to Capture			
8563	2020	All other reportable incidents	35223	Endocardial/interventricular septal pacing lead	CT1684	Cardiac pacemakers and associated devices	E0604	Cardiac Perforation	A070102	High Capture Threshold			
									A24	No Apparent Device Problem			
							E0605	Cardiac Tamponade	A070102	High Capture Threshold			
									A24	No Apparent Device Problem			
							E0619	Pericardial Effusion	A070102	High Capture Threshold			
									A24	No Apparent Device Problem			
							E2321	Low Blood Pressure/ Hypotension	A070102	High Capture Threshold			
									A24	No Apparent Device Problem			
							E2403	No Clinical Signs, Symptoms or Conditions	A070102	High Capture Threshold			
									A24	No Apparent Device Problem			
							F1203	Life Threatening Illness or Injury	A070102	High Capture Threshold			
									A24	No Apparent Device Problem			
							F23	Unexpected Medical Intervention	A070102	High Capture Threshold			
									A24	No Apparent Device Problem			
8564	2019	All other reportable incidents	47265	Dual-chamber implantable pacemaker, rate-responsive	CT141	Electrodes/Leads and associated devices	F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery			
					CT145	Cardiovascular devices	F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery			
					CT1684	Cardiac pacemakers and associated devices	F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery			
					CT1685	Cardiac pacemakers	F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery			
					CT233	Surgical	F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery			
					CT2406	Implantable	F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery			
					CT2407	Active implantable	F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery			
					CT257	Implantable cardiac pacemakers	F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery			
					CT2727	Active implant devices/ancillaries	F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery			
					CT2750	Implantable cardiac electrical stimulators and associated devices	F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery			
					CT2866	Implantable cardiac electrical stimulators	F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery			
					CT2985	Active	F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery			
					CT3001	Implantable metal	F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery			
					CT3002	Implantable synthetic polymer	F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery			
					CT3006	Implantable inorganic material	F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery			
					CT314	Electrically powered	F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery			
					CT321	Long-term surgical invasive	F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery			
					CT449	Generators and associated devices	F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery			
					CT981	Single use	F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery			
					CT983	Surgical invasive	F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery			
8565	2020	All other reportable incidents	36241	Pacemaker/defibrillator lead	CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A072201	High Impedance			
8566	2020	All other reportable incidents	36241	Pacemaker/defibrillator lead	CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A072201	High Impedance			
8567	2019	All other reportable incidents	37265	Dual-chamber implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	22/01/2020	22/01/2020	
8568	2019	All other reportable incidents	35852	Single-chamber implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A12	Connection Problem	01/03/2019	01/03/2019	
8569	2019	All other reportable incidents	17882	Physiologic-monitoring defibrillation system	CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A12	Connection Problem			
8570	2019	All other reportable incidents	35852	Single-chamber implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A070909	Oversensing			
8571	2019	All other reportable incidents	37265	Dual-chamber implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A071301	Failure to Deliver Shock/Stimulation			
8572	2020	All other reportable incidents	47265	Dual-chamber implantable pacemaker, rate-responsive	CT141	Electrodes/Leads and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A0701	Capturing Problem	10/01/2020	10/01/2020	
							F1901	Additional Surgery	A0701	Capturing Problem	10/01/2020	10/01/2020	
							F1905	Device Revision or Replacement	A0701	Capturing Problem	10/01/2020	10/01/2020	
					CT145	Cardiovascular devices	E2403	No Clinical Signs, Symptoms or Conditions	A0701	Capturing Problem	10/01/2020	10/01/2020	
							F1901	Additional Surgery	A0701	Capturing Problem	10/01/2020	10/01/2020	
							F1905	Device Revision or Replacement	A0701	Capturing Problem	10/01/2020	10/01/2020	
					CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A0701	Capturing Problem	10/01/2020	10/01/2020	
							F1901	Additional Surgery	A0701	Capturing Problem	10/01/2020	10/01/2020	
							F1905	Device Revision or Replacement	A0701	Capturing Problem	10/01/2020	10/01/2020	
					CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A0701	Capturing Problem	10/01/2020	10/01/2020	
							F1901	Additional Surgery	A0701	Capturing Problem	10/01/2020	10/01/2020	
							F1905	Device Revision or Replacement	A0701	Capturing Problem	10/01/2020	10/01/2020	
					CT233	Surgical	E2403	No Clinical Signs, Symptoms or Conditions	A0701	Capturing Problem	10/01/2020	10/01/2020	
							F1901	Additional Surgery	A0701	Capturing Problem	10/01/2020	10/01/2020	

							F1905	Device Revision or Replacement	A0701	Capturing Problem	10/01/2020	10/01/2020
					CT2406	Implantable	E2403	No Clinical Signs, Symptoms or Conditions	A0701	Capturing Problem	10/01/2020	10/01/2020
							F1901	Additional Surgery	A0701	Capturing Problem	10/01/2020	10/01/2020
							F1905	Device Revision or Replacement	A0701	Capturing Problem	10/01/2020	10/01/2020
					CT2407	Active implantable	E2403	No Clinical Signs, Symptoms or Conditions	A0701	Capturing Problem	10/01/2020	10/01/2020
							F1901	Additional Surgery	A0701	Capturing Problem	10/01/2020	10/01/2020
							F1905	Device Revision or Replacement	A0701	Capturing Problem	10/01/2020	10/01/2020
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A0701	Capturing Problem	10/01/2020	10/01/2020
							F1901	Additional Surgery	A0701	Capturing Problem	10/01/2020	10/01/2020
							F1905	Device Revision or Replacement	A0701	Capturing Problem	10/01/2020	10/01/2020
					CT2727	Active implant devices/ancillaries	E2403	No Clinical Signs, Symptoms or Conditions	A0701	Capturing Problem	10/01/2020	10/01/2020
							F1901	Additional Surgery	A0701	Capturing Problem	10/01/2020	10/01/2020
							F1905	Device Revision or Replacement	A0701	Capturing Problem	10/01/2020	10/01/2020
					CT2750	Implantable cardiac electrical stimulators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A0701	Capturing Problem	10/01/2020	10/01/2020
							F1901	Additional Surgery	A0701	Capturing Problem	10/01/2020	10/01/2020
							F1905	Device Revision or Replacement	A0701	Capturing Problem	10/01/2020	10/01/2020
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A0701	Capturing Problem	10/01/2020	10/01/2020
							F1901	Additional Surgery	A0701	Capturing Problem	10/01/2020	10/01/2020
							F1905	Device Revision or Replacement	A0701	Capturing Problem	10/01/2020	10/01/2020
					CT2985	Active	E2403	No Clinical Signs, Symptoms or Conditions	A0701	Capturing Problem	10/01/2020	10/01/2020
							F1901	Additional Surgery	A0701	Capturing Problem	10/01/2020	10/01/2020
							F1905	Device Revision or Replacement	A0701	Capturing Problem	10/01/2020	10/01/2020
					CT3001	Implantable metal	E2403	No Clinical Signs, Symptoms or Conditions	A0701	Capturing Problem	10/01/2020	10/01/2020
							F1901	Additional Surgery	A0701	Capturing Problem	10/01/2020	10/01/2020
							F1905	Device Revision or Replacement	A0701	Capturing Problem	10/01/2020	10/01/2020
					CT3002	Implantable synthetic polymer	E2403	No Clinical Signs, Symptoms or Conditions	A0701	Capturing Problem	10/01/2020	10/01/2020
							F1901	Additional Surgery	A0701	Capturing Problem	10/01/2020	10/01/2020
							F1905	Device Revision or Replacement	A0701	Capturing Problem	10/01/2020	10/01/2020
					CT3006	Implantable inorganic material	E2403	No Clinical Signs, Symptoms or Conditions	A0701	Capturing Problem	10/01/2020	10/01/2020
							F1901	Additional Surgery	A0701	Capturing Problem	10/01/2020	10/01/2020
							F1905	Device Revision or Replacement	A0701	Capturing Problem	10/01/2020	10/01/2020
					CT314	Electrically powered	E2403	No Clinical Signs, Symptoms or Conditions	A0701	Capturing Problem	10/01/2020	10/01/2020
							F1901	Additional Surgery	A0701	Capturing Problem	10/01/2020	10/01/2020
							F1905	Device Revision or Replacement	A0701	Capturing Problem	10/01/2020	10/01/2020
					CT321	Long-term surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A0701	Capturing Problem	10/01/2020	10/01/2020
							F1901	Additional Surgery	A0701	Capturing Problem	10/01/2020	10/01/2020
							F1905	Device Revision or Replacement	A0701	Capturing Problem	10/01/2020	10/01/2020
					CT449	Generators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A0701	Capturing Problem	10/01/2020	10/01/2020
							F1901	Additional Surgery	A0701	Capturing Problem	10/01/2020	10/01/2020
							F1905	Device Revision or Replacement	A0701	Capturing Problem	10/01/2020	10/01/2020
					CT981	Single use	E2403	No Clinical Signs, Symptoms or Conditions	A0701	Capturing Problem	10/01/2020	10/01/2020
							F1901	Additional Surgery	A0701	Capturing Problem	10/01/2020	10/01/2020
							F1905	Device Revision or Replacement	A0701	Capturing Problem	10/01/2020	10/01/2020
					CT983	Surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A0701	Capturing Problem	10/01/2020	10/01/2020
							F1901	Additional Surgery	A0701	Capturing Problem	10/01/2020	10/01/2020
							F1905	Device Revision or Replacement	A0701	Capturing Problem	10/01/2020	10/01/2020
8573	2019	All other reportable incidents	47270	Cardiac resynchronization therapy implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	08/03/2019	
8574	2020	All other reportable incidents	35852	Single-chamber implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A0712	Pacing Problem		
8575	2020	All other reportable incidents	35852	Single-chamber implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A070909	Oversensing		
8576	2019	All other reportable incidents	47265	Dual-chamber implantable pacemaker, rate-responsive	CT141	Electrodes/Leads and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
					CT145	Cardiovascular devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
					CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
					CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
					CT233	Surgical	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
					CT2406	Implantable	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
					CT2407	Active implantable	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
					CT2727	Active implant devices/ancillaries	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
					CT2750	Implantable cardiac electrical stimulators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
					CT2985	Active	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
					CT3001	Implantable metal	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
					CT3002	Implantable synthetic polymer	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
					CT3006	Implantable inorganic material	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
					CT314	Electrically powered	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
					CT321	Long-term surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
					CT449	Generators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
					CT981	Single use	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
					CT983	Surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
8577	2019	All other reportable incidents	37265	Dual-chamber implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A070504	Premature Discharge of Battery	02/10/2018	
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	02/10/2018	
8578	2019	All other reportable incidents	47266	Single-chamber implantable pacemaker, fixed-rate/demand	CT141	Electrodes/Leads and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
					CT145	Cardiovascular devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
					CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
					CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
					CT233	Surgical	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
					CT2406	Implantable	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
					CT2407	Active implantable	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
					CT2727	Active implant devices/ancillaries	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
					CT2750	Implantable cardiac electrical stimulators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
					CT2985	Active	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
					CT3001	Implantable metal	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
					CT3002	Implantable synthetic polymer	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		

					CT3006	Implantable inorganic material	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery			
					CT314	Electrically powered	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery			
					CT321	Long-term surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery			
					CT449	Generators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery			
					CT981	Single use	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery			
					CT983	Surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery			
8579	2019	All other reportable incidents	47265	Dual-chamber implantable pacemaker, rate-responsive	CT141	Electrodes/Leads and associated devices	F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	01/03/2019		
					CT145	Cardiovascular devices	F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	01/03/2019		
					CT1684	Cardiac pacemakers and associated devices	F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	01/03/2019		
					CT1685	Cardiac pacemakers	F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	01/03/2019		
					CT233	Surgical	F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	01/03/2019		
					CT2406	Implantable	F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	01/03/2019		
					CT2407	Active implantable	F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	01/03/2019		
					CT257	Implantable cardiac pacemakers	F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	01/03/2019		
					CT2727	Active implant devices/ancillaries	F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	01/03/2019		
					CT2750	Implantable cardiac electrical stimulators and associated devices	F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	01/03/2019		
					CT2866	Implantable cardiac electrical stimulators	F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	01/03/2019		
					CT2985	Active	F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	01/03/2019		
					CT3001	Implantable metal	F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	01/03/2019		
					CT3002	Implantable synthetic polymer	F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	01/03/2019		
					CT3006	Implantable inorganic material	F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	01/03/2019		
					CT314	Electrically powered	F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	01/03/2019		
					CT321	Long-term surgical invasive	F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	01/03/2019		
					CT449	Generators and associated devices	F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	01/03/2019		
					CT981	Single use	F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	01/03/2019		
					CT983	Surgical invasive	F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	01/03/2019		
8580	2019	All other reportable incidents	36241	Pacemaker/defibrillator lead	CT1684	Cardiac pacemakers and associated devices	E2104	Electric Shock	A071302	Inappropriate/Inadequate Shock/Stimulation			
8581	2019	All other reportable incidents	37265	Dual-chamber implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery			
8582	2019	All other reportable incidents	17882	Physiologic-monitoring defibrillation system	CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070803	Failure to Power Up			
8583	2019	All other reportable incidents	47270	Cardiac resynchronization therapy implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E2104	Electric Shock	A070909	Oversensing			
8584	2020	All other reportable incidents	47263	Cardiac resynchronization therapy implantable pacemaker	CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery			
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery			
					CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery			
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery			
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery			
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery			
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery			
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery			
8585	2020	All other reportable incidents	36241	Pacemaker/defibrillator lead	CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A051201	Device Dislodged or Dislocated	31/12/2019	31/12/2019	
									A070101	Failure to Capture	31/12/2019	31/12/2019	
									A070904	Failure to Select Signal	31/12/2019	31/12/2019	
							F19	Surgical Intervention	A051201	Device Dislodged or Dislocated	31/12/2019	31/12/2019	
									A070101	Failure to Capture	31/12/2019	31/12/2019	
									A070904	Failure to Select Signal	31/12/2019	31/12/2019	
							F1905	Device Revision or Replacement	A051201	Device Dislodged or Dislocated	31/12/2019	31/12/2019	
									A070101	Failure to Capture	31/12/2019	31/12/2019	
									A070904	Failure to Select Signal	31/12/2019	31/12/2019	
8586	2020	All other reportable incidents	47265	Dual-chamber implantable pacemaker, rate-responsive	CT141	Electrodes/Leads and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A072201	High Impedance			
					CT145	Cardiovascular devices	E2403	No Clinical Signs, Symptoms or Conditions	A072201	High Impedance			
					CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A072201	High Impedance			
					CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A072201	High Impedance			
					CT233	Surgical	E2403	No Clinical Signs, Symptoms or Conditions	A072201	High Impedance			
					CT2406	Implantable	E2403	No Clinical Signs, Symptoms or Conditions	A072201	High Impedance			
					CT2407	Active implantable	E2403	No Clinical Signs, Symptoms or Conditions	A072201	High Impedance			
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A072201	High Impedance			
					CT2727	Active implant devices/ancillaries	E2403	No Clinical Signs, Symptoms or Conditions	A072201	High Impedance			
					CT2750	Implantable cardiac electrical stimulators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A072201	High Impedance			
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A072201	High Impedance			
					CT2985	Active	E2403	No Clinical Signs, Symptoms or Conditions	A072201	High Impedance			
					CT3001	Implantable metal	E2403	No Clinical Signs, Symptoms or Conditions	A072201	High Impedance			
					CT3002	Implantable synthetic polymer	E2403	No Clinical Signs, Symptoms or Conditions	A072201	High Impedance			
					CT3006	Implantable inorganic material	E2403	No Clinical Signs, Symptoms or Conditions	A072201	High Impedance			
					CT314	Electrically powered	E2403	No Clinical Signs, Symptoms or Conditions	A072201	High Impedance			
					CT321	Long-term surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A072201	High Impedance			
					CT449	Generators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A072201	High Impedance			
					CT981	Single use	E2403	No Clinical Signs, Symptoms or Conditions	A072201	High Impedance			
					CT983	Surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A072201	High Impedance			
8587	2019	All other reportable incidents	47265	Dual-chamber implantable pacemaker, rate-responsive	CT141	Electrodes/Leads and associated devices	F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A0721	Circuit Failure	28/02/2019		
							F1905	Device Revision or Replacement	A0721	Circuit Failure	28/02/2019		
					CT145	Cardiovascular devices	F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A0721	Circuit Failure	28/02/2019		
							F1905	Device Revision or Replacement	A0721	Circuit Failure	28/02/2019		
					CT1684	Cardiac pacemakers and associated devices	F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A0721	Circuit Failure	28/02/2019		
							F1905	Device Revision or Replacement	A0721	Circuit Failure	28/02/2019		
					CT1685	Cardiac pacemakers	F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A0721	Circuit Failure	28/02/2019		
							F1905	Device Revision or Replacement	A0721	Circuit Failure	28/02/2019		
					CT233	Surgical	F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A0721	Circuit Failure	28/02/2019		
							F1905	Device Revision or Replacement	A0721	Circuit Failure	28/02/2019		
					CT2406	Implantable	F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A0721	Circuit Failure	28/02/2019		
							F1905	Device Revision or Replacement	A0721	Circuit Failure	28/02/2019		
					CT2407	Active implantable	F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A0721	Circuit Failure	28/02/2019		
							F1905	Device Revision or Replacement	A0721	Circuit Failure	28/02/2019		
					CT257	Implantable cardiac pacemakers	F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A0721	Circuit Failure	28/02/2019		
							F1905	Device Revision or Replacement	A0721	Circuit Failure	28/02/2019		
					CT2727	Active implant devices/ancillaries	F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A0721	Circuit Failure	28/02/2019		
							F1905	Device Revision or Replacement	A0721	Circuit Failure	28/02/2019		

					CT2750	Implantable cardiac electrical stimulators and associated devices	F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A0721	Circuit Failure	28/02/2019		
							F1905	Device Revision or Replacement	A0721	Circuit Failure	28/02/2019		
					CT2866	Implantable cardiac electrical stimulators	F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A0721	Circuit Failure	28/02/2019		
							F1905	Device Revision or Replacement	A0721	Circuit Failure	28/02/2019		
					CT2985	Active	F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A0721	Circuit Failure	28/02/2019		
							F1905	Device Revision or Replacement	A0721	Circuit Failure	28/02/2019		
					CT3001	Implantable metal	F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A0721	Circuit Failure	28/02/2019		
							F1905	Device Revision or Replacement	A0721	Circuit Failure	28/02/2019		
					CT3002	Implantable synthetic polymer	F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A0721	Circuit Failure	28/02/2019		
							F1905	Device Revision or Replacement	A0721	Circuit Failure	28/02/2019		
					CT3006	Implantable inorganic material	F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A0721	Circuit Failure	28/02/2019		
							F1905	Device Revision or Replacement	A0721	Circuit Failure	28/02/2019		
					CT314	Electrically powered	F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A0721	Circuit Failure	28/02/2019		
							F1905	Device Revision or Replacement	A0721	Circuit Failure	28/02/2019		
					CT321	Long-term surgical invasive	F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A0721	Circuit Failure	28/02/2019		
							F1905	Device Revision or Replacement	A0721	Circuit Failure	28/02/2019		
					CT449	Generators and associated devices	F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A0721	Circuit Failure	28/02/2019		
							F1905	Device Revision or Replacement	A0721	Circuit Failure	28/02/2019		
					CT981	Single use	F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A0721	Circuit Failure	28/02/2019		
							F1905	Device Revision or Replacement	A0721	Circuit Failure	28/02/2019		
					CT983	Surgical invasive	F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A0721	Circuit Failure	28/02/2019		
							F1905	Device Revision or Replacement	A0721	Circuit Failure	28/02/2019		
8588	2019	All other reportable incidents	47265	Dual-chamber implantable pacemaker, rate-responsive	CT141	Electrodes/Leads and associated devices	F1905	Device Revision or Replacement	A0712	Pacing Problem	05/03/2019		
					CT145	Cardiovascular devices	F1905	Device Revision or Replacement	A0712	Pacing Problem	05/03/2019		
					CT1684	Cardiac pacemakers and associated devices	F1905	Device Revision or Replacement	A0712	Pacing Problem	05/03/2019		
					CT1685	Cardiac pacemakers	F1905	Device Revision or Replacement	A0712	Pacing Problem	05/03/2019		
					CT233	Surgical	F1905	Device Revision or Replacement	A0712	Pacing Problem	05/03/2019		
					CT2406	Implantable	F1905	Device Revision or Replacement	A0712	Pacing Problem	05/03/2019		
					CT2407	Active implantable	F1905	Device Revision or Replacement	A0712	Pacing Problem	05/03/2019		
					CT257	Implantable cardiac pacemakers	F1905	Device Revision or Replacement	A0712	Pacing Problem	05/03/2019		
					CT2727	Active implant devices/ancillaries	F1905	Device Revision or Replacement	A0712	Pacing Problem	05/03/2019		
					CT2750	Implantable cardiac electrical stimulators and associated devices	F1905	Device Revision or Replacement	A0712	Pacing Problem	05/03/2019		
					CT2866	Implantable cardiac electrical stimulators	F1905	Device Revision or Replacement	A0712	Pacing Problem	05/03/2019		
					CT2985	Active	F1905	Device Revision or Replacement	A0712	Pacing Problem	05/03/2019		
					CT3001	Implantable metal	F1905	Device Revision or Replacement	A0712	Pacing Problem	05/03/2019		
					CT3002	Implantable synthetic polymer	F1905	Device Revision or Replacement	A0712	Pacing Problem	05/03/2019		
					CT3006	Implantable inorganic material	F1905	Device Revision or Replacement	A0712	Pacing Problem	05/03/2019		
					CT314	Electrically powered	F1905	Device Revision or Replacement	A0712	Pacing Problem	05/03/2019		
					CT321	Long-term surgical invasive	F1905	Device Revision or Replacement	A0712	Pacing Problem	05/03/2019		
					CT449	Generators and associated devices	F1905	Device Revision or Replacement	A0712	Pacing Problem	05/03/2019		
					CT981	Single use	F1905	Device Revision or Replacement	A0712	Pacing Problem	05/03/2019		
					CT983	Surgical invasive	F1905	Device Revision or Replacement	A0712	Pacing Problem	05/03/2019		
8589	2019	All other reportable incidents	47265	Dual-chamber implantable pacemaker, rate-responsive	CT141	Electrodes/Leads and associated devices	F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	01/03/2019		
					CT145	Cardiovascular devices	F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	01/03/2019		
					CT1684	Cardiac pacemakers and associated devices	F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	01/03/2019		
					CT1685	Cardiac pacemakers	F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	01/03/2019		
					CT233	Surgical	F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	01/03/2019		
					CT2406	Implantable	F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	01/03/2019		
					CT2407	Active implantable	F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	01/03/2019		
					CT257	Implantable cardiac pacemakers	F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	01/03/2019		
					CT2727	Active implant devices/ancillaries	F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	01/03/2019		
					CT2750	Implantable cardiac electrical stimulators and associated devices	F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	01/03/2019		
					CT2866	Implantable cardiac electrical stimulators	F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	01/03/2019		
					CT2985	Active	F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	01/03/2019		
					CT3001	Implantable metal	F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	01/03/2019		
					CT3002	Implantable synthetic polymer	F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	01/03/2019		
					CT3006	Implantable inorganic material	F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	01/03/2019		
					CT314	Electrically powered	F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	01/03/2019		
					CT321	Long-term surgical invasive	F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	01/03/2019		
					CT449	Generators and associated devices	F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	01/03/2019		
					CT981	Single use	F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	01/03/2019		
					CT983	Surgical invasive	F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	01/03/2019		
8590	2019	All other reportable incidents	36241	Pacemaker/defibrillator lead	CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A0510	Retraction Problem	05/03/2019		
8591	2020	All other reportable incidents	47263	Cardiac resynchronization therapy implantable pacemaker	CT1684	Cardiac pacemakers and associated devices	F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	23/01/2020	23/01/2020	
					CT1685	Cardiac pacemakers	F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	23/01/2020	23/01/2020	
					CT257	Implantable cardiac pacemakers	F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	23/01/2020	23/01/2020	
					CT2866	Implantable cardiac electrical stimulators	F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	23/01/2020	23/01/2020	
8592	2019	All other reportable incidents	17882	Physiologic-monitoring defibrillation system	CT1684	Cardiac pacemakers and associated devices	F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A0902	Display or Visual Feedback Problem			
8593	2020	All other reportable incidents	37265	Dual-chamber implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A072201	High Impedance			
							F26	No Health Consequences or Impact	A072201	High Impedance			
8594	2020	All other reportable incidents	47265	Dual-chamber implantable pacemaker, rate-responsive	CT141	Electrodes/Leads and associated devices	F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	02/01/2020	02/01/2020	
					CT145	Cardiovascular devices	F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	02/01/2020	02/01/2020	
					CT1684	Cardiac pacemakers and associated devices	F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	02/01/2020	02/01/2020	
					CT1685	Cardiac pacemakers	F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	02/01/2020	02/01/2020	
					CT233	Surgical	F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	02/01/2020	02/01/2020	
					CT2406	Implantable	F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	02/01/2020	02/01/2020	
					CT2407	Active implantable	F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	02/01/2020	02/01/2020	
					CT257	Implantable cardiac pacemakers	F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	02/01/2020	02/01/2020	
					CT2727	Active implant devices/ancillaries	F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	02/01/2020	02/01/2020	
					CT2750	Implantable cardiac electrical stimulators and associated devices	F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	02/01/2020	02/01/2020	
					CT2866	Implantable cardiac electrical stimulators	F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	02/01/2020	02/01/2020	
					CT2985	Active	F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	02/01/2020	02/01/2020	
					CT3001	Implantable metal	F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	02/01/2020	02/01/2020	
					CT3002	Implantable synthetic polymer	F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	02/01/2020	02/01/2020	
					CT3006	Implantable inorganic material	F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	02/01/2020	02/01/2020	

					CT314	Electrically powered	F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	02/01/2020	02/01/2020
					CT321	Long-term surgical invasive	F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	02/01/2020	02/01/2020
					CT449	Generators and associated devices	F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	02/01/2020	02/01/2020
					CT981	Single use	F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	02/01/2020	02/01/2020
					CT983	Surgical invasive	F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	02/01/2020	02/01/2020
8595	2020	All other reportable incidents	35852	Single-chamber implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E2104	Electric Shock	A071302	Inappropriate/Inadequate Shock/Stimulation		
8596	2019	All other reportable incidents	47263	Cardiac resynchronization therapy implantable pacemaker	CT1684	Cardiac pacemakers and associated devices	F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	07/03/2019	
					CT1685	Cardiac pacemakers	F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	07/03/2019	
					CT257	Implantable cardiac pacemakers	F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	07/03/2019	
					CT2866	Implantable cardiac electrical stimulators	F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	07/03/2019	
8597	2020	All other reportable incidents	37265	Dual-chamber implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	F1905	Device Revision or Replacement	A090201	Device Displays Incorrect Message	07/01/2020	07/01/2020
8598	2020	All other reportable incidents	37265	Dual-chamber implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	F1905	Device Revision or Replacement	A090201	Device Displays Incorrect Message	07/01/2020	07/01/2020
8599	2020	All other reportable incidents	47270	Cardiac resynchronization therapy implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A0712	Pacing Problem		
8600	2020	Death	35223	Endocardial/interventricular septal pacing lead	CT1684	Cardiac pacemakers and associated devices	E2401	Insufficient Information	A24	No Apparent Device Problem		
8601	2019	All other reportable incidents	47265	Dual-chamber implantable pacemaker, rate-responsive	CT141	Electrodes/Leads and associated devices	F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	21/02/2019	
					CT145	Cardiovascular devices	F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	21/02/2019	
					CT1684	Cardiac pacemakers and associated devices	F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	21/02/2019	
					CT1685	Cardiac pacemakers	F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	21/02/2019	
					CT233	Surgical	F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	21/02/2019	
					CT2406	Implantable	F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	21/02/2019	
					CT2407	Active implantable	F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	21/02/2019	
					CT257	Implantable cardiac pacemakers	F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	21/02/2019	
					CT2727	Active implant devices/ancillaries	F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	21/02/2019	
					CT2750	Implantable cardiac electrical stimulators and associated devices	F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	21/02/2019	
					CT2866	Implantable cardiac electrical stimulators	F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	21/02/2019	
					CT2985	Active	F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	21/02/2019	
					CT3001	Implantable metal	F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	21/02/2019	
					CT3002	Implantable synthetic polymer	F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	21/02/2019	
					CT3006	Implantable inorganic material	F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	21/02/2019	
					CT314	Electrically powered	F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	21/02/2019	
					CT321	Long-term surgical invasive	F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	21/02/2019	
					CT449	Generators and associated devices	F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	21/02/2019	
					CT981	Single use	F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	21/02/2019	
					CT983	Surgical invasive	F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	21/02/2019	
8602	2019	All other reportable incidents	35852	Single-chamber implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A070909	Oversensing		
8603	2019	All other reportable incidents	35223	Endocardial/interventricular septal pacing lead	CT1684	Cardiac pacemakers and associated devices	F14	Prolonged Episode of Care	A072201	High Impedance		
8604	2020	All other reportable incidents	36241	Pacemaker/defibrillator lead	CT1684	Cardiac pacemakers and associated devices	F1905	Device Revision or Replacement	A051201	Device Dislodged or Dislocated		
8605	2020	All other reportable incidents	36241	Pacemaker/defibrillator lead	CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A072201	High Impedance		
8606	2020	All other reportable incidents	35852	Single-chamber implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A071302	Inappropriate/Inadequate Shock/Stimulation		
8607	2020	All other reportable incidents	35852	Single-chamber implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A0711	Interrogation Problem		
8608	2020	All other reportable incidents	36241	Pacemaker/defibrillator lead	CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A040101	Fracture		
							F26	No Health Consequences or Impact	A040101	Fracture		
8609	2019	All other reportable incidents	36241	Pacemaker/defibrillator lead	CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A0709	Device Sensing Problem		
8610	2019	All other reportable incidents	37265	Dual-chamber implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A070910	Undersensing		
8611	2019	All other reportable incidents	47265	Dual-chamber implantable pacemaker, rate-responsive	CT141	Electrodes/Leads and associated devices	F1905	Device Revision or Replacement	A071208	Pacing Intermittently	22/02/2019	
					CT145	Cardiovascular devices	F1905	Device Revision or Replacement	A071208	Pacing Intermittently	22/02/2019	
					CT1684	Cardiac pacemakers and associated devices	F1905	Device Revision or Replacement	A071208	Pacing Intermittently	22/02/2019	
					CT1685	Cardiac pacemakers	F1905	Device Revision or Replacement	A071208	Pacing Intermittently	22/02/2019	
					CT233	Surgical	F1905	Device Revision or Replacement	A071208	Pacing Intermittently	22/02/2019	
					CT2406	Implantable	F1905	Device Revision or Replacement	A071208	Pacing Intermittently	22/02/2019	
					CT2407	Active implantable	F1905	Device Revision or Replacement	A071208	Pacing Intermittently	22/02/2019	
					CT257	Implantable cardiac pacemakers	F1905	Device Revision or Replacement	A071208	Pacing Intermittently	22/02/2019	
					CT2727	Active implant devices/ancillaries	F1905	Device Revision or Replacement	A071208	Pacing Intermittently	22/02/2019	
					CT2750	Implantable cardiac electrical stimulators and associated devices	F1905	Device Revision or Replacement	A071208	Pacing Intermittently	22/02/2019	
					CT2866	Implantable cardiac electrical stimulators	F1905	Device Revision or Replacement	A071208	Pacing Intermittently	22/02/2019	
					CT2985	Active	F1905	Device Revision or Replacement	A071208	Pacing Intermittently	22/02/2019	
					CT3001	Implantable metal	F1905	Device Revision or Replacement	A071208	Pacing Intermittently	22/02/2019	
					CT3002	Implantable synthetic polymer	F1905	Device Revision or Replacement	A071208	Pacing Intermittently	22/02/2019	
					CT3006	Implantable inorganic material	F1905	Device Revision or Replacement	A071208	Pacing Intermittently	22/02/2019	
					CT314	Electrically powered	F1905	Device Revision or Replacement	A071208	Pacing Intermittently	22/02/2019	
					CT321	Long-term surgical invasive	F1905	Device Revision or Replacement	A071208	Pacing Intermittently	22/02/2019	
					CT449	Generators and associated devices	F1905	Device Revision or Replacement	A071208	Pacing Intermittently	22/02/2019	
					CT981	Single use	F1905	Device Revision or Replacement	A071208	Pacing Intermittently	22/02/2019	
					CT983	Surgical invasive	F1905	Device Revision or Replacement	A071208	Pacing Intermittently	22/02/2019	
8612	2019	All other reportable incidents	37265	Dual-chamber implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
8613	2019	All other reportable incidents	47265	Dual-chamber implantable pacemaker, rate-responsive	CT141	Electrodes/Leads and associated devices	F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery		
					CT145	Cardiovascular devices	F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery		
					CT1684	Cardiac pacemakers and associated devices	F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery		
					CT1685	Cardiac pacemakers	F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery		
					CT233	Surgical	F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery		
					CT2406	Implantable	F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery		
					CT2407	Active implantable	F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery		
					CT257	Implantable cardiac pacemakers	F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery		
					CT2727	Active implant devices/ancillaries	F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery		
					CT2750	Implantable cardiac electrical stimulators and associated devices	F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery		
					CT2866	Implantable cardiac electrical stimulators	F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery		
					CT2985	Active	F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery		
					CT3001	Implantable metal	F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery		
					CT3002	Implantable synthetic polymer	F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery		
					CT3006	Implantable inorganic material	F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery		
					CT314	Electrically powered	F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery		
					CT321	Long-term surgical invasive	F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery		
					CT449	Generators and associated devices	F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery		

					CT981	Single use	F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery		
					CT983	Surgical invasive	F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery		
8614	2019	All other reportable incidents	36241	Pacemaker/defibrillator lead	CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A0510	Retraction Problem	28/02/2019	
8615	2019	All other reportable incidents	37265	Dual-chamber implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A0709	Device Sensing Problem		
8616	2019	All other reportable incidents	36241	Pacemaker/defibrillator lead	CT1684	Cardiac pacemakers and associated devices	F1905	Device Revision or Replacement	A0506	Mechanical Jam	06/03/2019	
8617	2019	All other reportable incidents	37265	Dual-chamber implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	04/02/2019	
8618	2020	All other reportable incidents	35852	Single-chamber implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	17/01/2020	17/01/2020
							F1901	Additional Surgery	A070504	Premature Discharge of Battery	17/01/2020	17/01/2020
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	17/01/2020	17/01/2020
8619	2020	All other reportable incidents	35223	Endocardial/interventricular septal pacing lead	CT1684	Cardiac pacemakers and associated devices	E0604	Cardiac Perforation	A0701	Capturing Problem		
									A0709	Device Sensing Problem		
									A072202	Low Impedance		
									A24	No Apparent Device Problem		
							E2114	Perforation	A0701	Capturing Problem		
									A0709	Device Sensing Problem		
									A072202	Low Impedance		
									A24	No Apparent Device Problem		
							F12	Serious Injury/ Illness/ Impairment	A0701	Capturing Problem		
									A0709	Device Sensing Problem		
									A072202	Low Impedance		
									A24	No Apparent Device Problem		
8620	2019	All other reportable incidents	47265	Dual-chamber implantable pacemaker, rate-responsive	CT141	Electrodes/Leads and associated devices	F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	15/03/2019	
					CT145	Cardiovascular devices	F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	15/03/2019	
					CT1684	Cardiac pacemakers and associated devices	F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	15/03/2019	
					CT1685	Cardiac pacemakers	F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	15/03/2019	
					CT233	Surgical	F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	15/03/2019	
					CT2406	Implantable	F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	15/03/2019	
					CT2407	Active implantable	F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	15/03/2019	
					CT257	Implantable cardiac pacemakers	F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	15/03/2019	
					CT2727	Active implant devices/ancillaries	F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	15/03/2019	
					CT2750	Implantable cardiac electrical stimulators and associated devices	F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	15/03/2019	
					CT2866	Implantable cardiac electrical stimulators	F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	15/03/2019	
					CT2985	Active	F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	15/03/2019	
					CT3001	Implantable metal	F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	15/03/2019	
					CT3002	Implantable synthetic polymer	F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	15/03/2019	
					CT3006	Implantable inorganic material	F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	15/03/2019	
					CT314	Electrically powered	F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	15/03/2019	
					CT321	Long-term surgical invasive	F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	15/03/2019	
					CT449	Generators and associated devices	F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	15/03/2019	
					CT981	Single use	F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	15/03/2019	
					CT983	Surgical invasive	F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	15/03/2019	
8621	2019	All other reportable incidents	36241	Pacemaker/defibrillator lead	CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A010402	Migration		
8622	2019	All other reportable incidents	35223	Endocardial/interventricular septal pacing lead	CT1684	Cardiac pacemakers and associated devices	F1905	Device Revision or Replacement	A072201	High Impedance	30/01/2019	
8623	2019	All other reportable incidents	37265	Dual-chamber implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	29/03/2019	
8624	2020	All other reportable incidents	47265	Dual-chamber implantable pacemaker, rate-responsive	CT141	Electrodes/Leads and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070909	Oversensing		
							F26	No Health Consequences or Impact	A070909	Oversensing		
					CT145	Cardiovascular devices	E2403	No Clinical Signs, Symptoms or Conditions	A070909	Oversensing		
							F26	No Health Consequences or Impact	A070909	Oversensing		
					CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070909	Oversensing		
							F26	No Health Consequences or Impact	A070909	Oversensing		
					CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A070909	Oversensing		
							F26	No Health Consequences or Impact	A070909	Oversensing		
					CT233	Surgical	E2403	No Clinical Signs, Symptoms or Conditions	A070909	Oversensing		
							F26	No Health Consequences or Impact	A070909	Oversensing		
					CT2406	Implantable	E2403	No Clinical Signs, Symptoms or Conditions	A070909	Oversensing		
							F26	No Health Consequences or Impact	A070909	Oversensing		
					CT2407	Active implantable	E2403	No Clinical Signs, Symptoms or Conditions	A070909	Oversensing		
							F26	No Health Consequences or Impact	A070909	Oversensing		
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A070909	Oversensing		
							F26	No Health Consequences or Impact	A070909	Oversensing		
					CT2727	Active implant devices/ancillaries	E2403	No Clinical Signs, Symptoms or Conditions	A070909	Oversensing		
							F26	No Health Consequences or Impact	A070909	Oversensing		
					CT2750	Implantable cardiac electrical stimulators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070909	Oversensing		
							F26	No Health Consequences or Impact	A070909	Oversensing		
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A070909	Oversensing		
							F26	No Health Consequences or Impact	A070909	Oversensing		
					CT2985	Active	E2403	No Clinical Signs, Symptoms or Conditions	A070909	Oversensing		
							F26	No Health Consequences or Impact	A070909	Oversensing		
					CT3001	Implantable metal	E2403	No Clinical Signs, Symptoms or Conditions	A070909	Oversensing		
							F26	No Health Consequences or Impact	A070909	Oversensing		
					CT3002	Implantable synthetic polymer	E2403	No Clinical Signs, Symptoms or Conditions	A070909	Oversensing		
							F26	No Health Consequences or Impact	A070909	Oversensing		
					CT3006	Implantable inorganic material	E2403	No Clinical Signs, Symptoms or Conditions	A070909	Oversensing		
							F26	No Health Consequences or Impact	A070909	Oversensing		
					CT314	Electrically powered	E2403	No Clinical Signs, Symptoms or Conditions	A070909	Oversensing		
							F26	No Health Consequences or Impact	A070909	Oversensing		
					CT321	Long-term surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A070909	Oversensing		
							F26	No Health Consequences or Impact	A070909	Oversensing		
					CT449	Generators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070909	Oversensing		
							F26	No Health Consequences or Impact	A070909	Oversensing		
					CT981	Single use	E2403	No Clinical Signs, Symptoms or Conditions	A070909	Oversensing		
							F26	No Health Consequences or Impact	A070909	Oversensing		
					CT983	Surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A070909	Oversensing		
							F26	No Health Consequences or Impact	A070909	Oversensing		
8625	2020	All other reportable incidents	47267	Single-chamber implantable pacemaker, rate-responsive	CT141	Electrodes/Leads and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		

					CT145	Cardiovascular devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT233	Surgical	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT2406	Implantable	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT2407	Active implantable	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT2727	Active implant devices/ancillaries	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT2750	Implantable cardiac electrical stimulators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT2985	Active	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT3001	Implantable metal	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT3002	Implantable synthetic polymer	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT3006	Implantable inorganic material	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT314	Electrically powered	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT321	Long-term surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT449	Generators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT981	Single use	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT983	Surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
8626	2020	All other reportable incidents	47267	Single-chamber implantable pacemaker, rate-responsive	CT141	Electrodes/Leads and associated devices	F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	13/01/2020	13/01/2020
					CT145	Cardiovascular devices	F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	13/01/2020	13/01/2020
					CT1684	Cardiac pacemakers and associated devices	F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	13/01/2020	13/01/2020
					CT1685	Cardiac pacemakers	F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	13/01/2020	13/01/2020
					CT233	Surgical	F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	13/01/2020	13/01/2020
					CT2406	Implantable	F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	13/01/2020	13/01/2020
					CT2407	Active implantable	F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	13/01/2020	13/01/2020
					CT257	Implantable cardiac pacemakers	F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	13/01/2020	13/01/2020
					CT2727	Active implant devices/ancillaries	F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	13/01/2020	13/01/2020
					CT2750	Implantable cardiac electrical stimulators and associated devices	F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	13/01/2020	13/01/2020
					CT2866	Implantable cardiac electrical stimulators	F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	13/01/2020	13/01/2020
					CT2985	Active	F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	13/01/2020	13/01/2020
					CT3001	Implantable metal	F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	13/01/2020	13/01/2020
					CT3002	Implantable synthetic polymer	F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	13/01/2020	13/01/2020
					CT3006	Implantable inorganic material	F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	13/01/2020	13/01/2020
					CT314	Electrically powered	F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	13/01/2020	13/01/2020
					CT321	Long-term surgical invasive	F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	13/01/2020	13/01/2020
					CT449	Generators and associated devices	F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	13/01/2020	13/01/2020
					CT981	Single use	F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	13/01/2020	13/01/2020
					CT983	Surgical invasive	F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	13/01/2020	13/01/2020
8627	2020	All other reportable incidents	37265	Dual-chamber implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A071302	Inappropriate/Inadequate Shock/Stimulation		
							F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A071302	Inappropriate/Inadequate Shock/Stimulation		
							F26	No Health Consequences or Impact	A071302	Inappropriate/Inadequate Shock/Stimulation		
8628	2020	All other reportable incidents	47265	Dual-chamber implantable pacemaker, rate-responsive	CT141	Electrodes/Leads and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
					CT145	Cardiovascular devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
					CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
					CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
					CT233	Surgical	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
					CT2406	Implantable	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
					CT2407	Active implantable	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
					CT2727	Active implant devices/ancillaries	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
					CT2750	Implantable cardiac electrical stimulators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
					CT2985	Active	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
					CT3001	Implantable metal	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
					CT3002	Implantable synthetic polymer	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
					CT3006	Implantable inorganic material	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
					CT314	Electrically powered	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
					CT321	Long-term surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
					CT449	Generators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
					CT981	Single use	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
					CT983	Surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
8629	2019	All other reportable incidents	47265	Dual-chamber implantable pacemaker, rate-responsive	CT141	Electrodes/Leads and associated devices	F1905	Device Revision or Replacement	A1102	Application Program Problem	18/03/2019	
					CT145	Cardiovascular devices	F1905	Device Revision or Replacement	A1102	Application Program Problem	18/03/2019	
					CT1684	Cardiac pacemakers and associated devices	F1905	Device Revision or Replacement	A1102	Application Program Problem	18/03/2019	
					CT1685	Cardiac pacemakers	F1905	Device Revision or Replacement	A1102	Application Program Problem	18/03/2019	
					CT233	Surgical	F1905	Device Revision or Replacement	A1102	Application Program Problem	18/03/2019	

					CT2406	Implantable	F1905	Device Revision or Replacement	A1102	Application Program Problem	18/03/2019		
					CT2407	Active implantable	F1905	Device Revision or Replacement	A1102	Application Program Problem	18/03/2019		
					CT257	Implantable cardiac pacemakers	F1905	Device Revision or Replacement	A1102	Application Program Problem	18/03/2019		
					CT2727	Active implant devices/ancillaries	F1905	Device Revision or Replacement	A1102	Application Program Problem	18/03/2019		
					CT2750	Implantable cardiac electrical stimulators and associated devices	F1905	Device Revision or Replacement	A1102	Application Program Problem	18/03/2019		
					CT2866	Implantable cardiac electrical stimulators	F1905	Device Revision or Replacement	A1102	Application Program Problem	18/03/2019		
					CT2985	Active	F1905	Device Revision or Replacement	A1102	Application Program Problem	18/03/2019		
					CT3001	Implantable metal	F1905	Device Revision or Replacement	A1102	Application Program Problem	18/03/2019		
					CT3002	Implantable synthetic polymer	F1905	Device Revision or Replacement	A1102	Application Program Problem	18/03/2019		
					CT3006	Implantable inorganic material	F1905	Device Revision or Replacement	A1102	Application Program Problem	18/03/2019		
					CT314	Electrically powered	F1905	Device Revision or Replacement	A1102	Application Program Problem	18/03/2019		
					CT321	Long-term surgical invasive	F1905	Device Revision or Replacement	A1102	Application Program Problem	18/03/2019		
					CT449	Generators and associated devices	F1905	Device Revision or Replacement	A1102	Application Program Problem	18/03/2019		
					CT981	Single use	F1905	Device Revision or Replacement	A1102	Application Program Problem	18/03/2019		
					CT983	Surgical invasive	F1905	Device Revision or Replacement	A1102	Application Program Problem	18/03/2019		
8630	2019	All other reportable incidents	35852	Single-chamber implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A072201	High Impedance	01/05/2019	01/05/2019	
8631	2019	All other reportable incidents	36241	Pacemaker/defibrillator lead	CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A0104	Migration or Expulsion of Device			
8632	2020	All other reportable incidents	17882	Physiologic-monitoring defibrillation system	CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A071301	Failure to Deliver Shock/Stimulation			
							F27	Problem Identified During Non-Clinical Procedure	A071301	Failure to Deliver Shock/Stimulation			
8633	2019	All other reportable incidents	47265	Dual-chamber implantable pacemaker, rate-responsive	CT141	Electrodes/Leads and associated devices	F1905	Device Revision or Replacement	A0721	Circuit Failure	11/03/2019		
					CT145	Cardiovascular devices	F1905	Device Revision or Replacement	A0721	Circuit Failure	11/03/2019		
					CT1684	Cardiac pacemakers and associated devices	F1905	Device Revision or Replacement	A0721	Circuit Failure	11/03/2019		
					CT1685	Cardiac pacemakers	F1905	Device Revision or Replacement	A0721	Circuit Failure	11/03/2019		
					CT233	Surgical	F1905	Device Revision or Replacement	A0721	Circuit Failure	11/03/2019		
					CT2406	Implantable	F1905	Device Revision or Replacement	A0721	Circuit Failure	11/03/2019		
					CT2407	Active implantable	F1905	Device Revision or Replacement	A0721	Circuit Failure	11/03/2019		
					CT257	Implantable cardiac pacemakers	F1905	Device Revision or Replacement	A0721	Circuit Failure	11/03/2019		
					CT2727	Active implant devices/ancillaries	F1905	Device Revision or Replacement	A0721	Circuit Failure	11/03/2019		
					CT2750	Implantable cardiac electrical stimulators and associated devices	F1905	Device Revision or Replacement	A0721	Circuit Failure	11/03/2019		
					CT2866	Implantable cardiac electrical stimulators	F1905	Device Revision or Replacement	A0721	Circuit Failure	11/03/2019		
					CT2985	Active	F1905	Device Revision or Replacement	A0721	Circuit Failure	11/03/2019		
					CT3001	Implantable metal	F1905	Device Revision or Replacement	A0721	Circuit Failure	11/03/2019		
					CT3002	Implantable synthetic polymer	F1905	Device Revision or Replacement	A0721	Circuit Failure	11/03/2019		
					CT3006	Implantable inorganic material	F1905	Device Revision or Replacement	A0721	Circuit Failure	11/03/2019		
					CT314	Electrically powered	F1905	Device Revision or Replacement	A0721	Circuit Failure	11/03/2019		
					CT321	Long-term surgical invasive	F1905	Device Revision or Replacement	A0721	Circuit Failure	11/03/2019		
					CT449	Generators and associated devices	F1905	Device Revision or Replacement	A0721	Circuit Failure	11/03/2019		
					CT981	Single use	F1905	Device Revision or Replacement	A0721	Circuit Failure	11/03/2019		
					CT983	Surgical invasive	F1905	Device Revision or Replacement	A0721	Circuit Failure	11/03/2019		
8634	2019	All other reportable incidents	35223	Endocardial/interventricular septal pacing lead	CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A040101	Fracture			
8635	2019	All other reportable incidents	35852	Single-chamber implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A070909	Oversensing			
8636	2019	All other reportable incidents	17882	Physiologic-monitoring defibrillation system	CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A0907	No Device Output			
8637	2020	All other reportable incidents	17882	Physiologic-monitoring defibrillation system	CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070803	Failure to Power Up			
							F27	Problem Identified During Non-Clinical Procedure	A070803	Failure to Power Up			
8638	2019	All other reportable incidents	35852	Single-chamber implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem			
8639	2019	All other reportable incidents	47270	Cardiac resynchronization therapy implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	F1905	Device Revision or Replacement	A0705	Battery Problem	20/03/2019		
8640	2019	All other reportable incidents	47265	Dual-chamber implantable pacemaker, rate-responsive	CT141	Electrodes/Leads and associated devices	F1905	Device Revision or Replacement	A110204	Application Program Problem: Parameter Calculation Error	18/03/2019		
					CT145	Cardiovascular devices	F1905	Device Revision or Replacement	A110204	Application Program Problem: Parameter Calculation Error	18/03/2019		
					CT1684	Cardiac pacemakers and associated devices	F1905	Device Revision or Replacement	A110204	Application Program Problem: Parameter Calculation Error	18/03/2019		
					CT1685	Cardiac pacemakers	F1905	Device Revision or Replacement	A110204	Application Program Problem: Parameter Calculation Error	18/03/2019		
					CT233	Surgical	F1905	Device Revision or Replacement	A110204	Application Program Problem: Parameter Calculation Error	18/03/2019		
					CT2406	Implantable	F1905	Device Revision or Replacement	A110204	Application Program Problem: Parameter Calculation Error	18/03/2019		
					CT2407	Active implantable	F1905	Device Revision or Replacement	A110204	Application Program Problem: Parameter Calculation Error	18/03/2019		
					CT257	Implantable cardiac pacemakers	F1905	Device Revision or Replacement	A110204	Application Program Problem: Parameter Calculation Error	18/03/2019		
					CT2727	Active implant devices/ancillaries	F1905	Device Revision or Replacement	A110204	Application Program Problem: Parameter Calculation Error	18/03/2019		
					CT2750	Implantable cardiac electrical stimulators and associated devices	F1905	Device Revision or Replacement	A110204	Application Program Problem: Parameter Calculation Error	18/03/2019		
					CT2866	Implantable cardiac electrical stimulators	F1905	Device Revision or Replacement	A110204	Application Program Problem: Parameter Calculation Error	18/03/2019		
					CT2985	Active	F1905	Device Revision or Replacement	A110204	Application Program Problem: Parameter Calculation Error	18/03/2019		
					CT3001	Implantable metal	F1905	Device Revision or Replacement	A110204	Application Program Problem: Parameter Calculation Error	18/03/2019		
					CT3002	Implantable synthetic polymer	F1905	Device Revision or Replacement	A110204	Application Program Problem: Parameter Calculation Error	18/03/2019		
					CT3006	Implantable inorganic material	F1905	Device Revision or Replacement	A110204	Application Program Problem: Parameter Calculation Error	18/03/2019		
					CT314	Electrically powered	F1905	Device Revision or Replacement	A110204	Application Program Problem: Parameter Calculation Error	18/03/2019		
					CT321	Long-term surgical invasive	F1905	Device Revision or Replacement	A110204	Application Program Problem: Parameter Calculation Error	18/03/2019		
					CT449	Generators and associated devices	F1905	Device Revision or Replacement	A110204	Application Program Problem: Parameter Calculation Error	18/03/2019		
					CT981	Single use	F1905	Device Revision or Replacement	A110204	Application Program Problem: Parameter Calculation Error	18/03/2019		
					CT983	Surgical invasive	F1905	Device Revision or Replacement	A110204	Application Program Problem: Parameter Calculation Error	18/03/2019		
8641	2019	All other reportable incidents	35223	Endocardial/interventricular septal pacing lead	CT1684	Cardiac pacemakers and associated devices	F1905	Device Revision or Replacement	A040101	Fracture			
8642	2019	All other reportable incidents	35852	Single-chamber implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A070909	Oversensing			
8643	2020	All other reportable incidents	35852	Single-chamber implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	28/01/2020	28/01/2020	

							F1901	Additional Surgery	A070504	Premature Discharge of Battery	28/01/2020	28/01/2020
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	28/01/2020	28/01/2020
8644	2019	All other reportable incidents	47263	Cardiac resynchronization therapy implantable pacemaker	CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
					CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
8645	2020	All other reportable incidents	37265	Dual-chamber implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A071302	Inappropriate/Inadequate Shock/Stimulation		
8646	2020	All other reportable incidents	36241	Pacemaker/defibrillator lead	CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A072201	High Impedance		
							F26	No Health Consequences or Impact	A072201	High Impedance		
8647	2020	All other reportable incidents	47265	Dual-chamber implantable pacemaker, rate-responsive	CT141	Electrodes/Leads and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
					CT145	Cardiovascular devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
					CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
					CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
					CT233	Surgical	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
					CT2406	Implantable	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
					CT2407	Active implantable	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
					CT2727	Active implant devices/ancillaries	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
					CT2750	Implantable cardiac electrical stimulators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
					CT2985	Active	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
					CT3001	Implantable metal	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
					CT3002	Implantable synthetic polymer	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
					CT3006	Implantable inorganic material	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
					CT314	Electrically powered	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
					CT321	Long-term surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
					CT449	Generators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
					CT981	Single use	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
					CT983	Surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
8648	2019	All other reportable incidents	37265	Dual-chamber implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A070909	Oversensing		
8649	2019	All other reportable incidents	37265	Dual-chamber implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A070909	Oversensing		
8650	2019	All other reportable incidents	37265	Dual-chamber implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery		
8651	2019	All other reportable incidents	47265	Dual-chamber implantable pacemaker, rate-responsive	CT141	Electrodes/Leads and associated devices	F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	22/03/2019	
					CT145	Cardiovascular devices	F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	22/03/2019	
					CT1684	Cardiac pacemakers and associated devices	F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	22/03/2019	
					CT1685	Cardiac pacemakers	F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	22/03/2019	
					CT233	Surgical	F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	22/03/2019	
					CT2406	Implantable	F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	22/03/2019	
					CT2407	Active implantable	F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	22/03/2019	
					CT257	Implantable cardiac pacemakers	F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	22/03/2019	
					CT2727	Active implant devices/ancillaries	F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	22/03/2019	
					CT2750	Implantable cardiac electrical stimulators and associated devices	F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	22/03/2019	
					CT2866	Implantable cardiac electrical stimulators	F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	22/03/2019	
					CT2985	Active	F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	22/03/2019	
					CT3001	Implantable metal	F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	22/03/2019	
					CT3002	Implantable synthetic polymer	F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	22/03/2019	
					CT3006	Implantable inorganic material	F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	22/03/2019	
					CT314	Electrically powered	F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	22/03/2019	
					CT321	Long-term surgical invasive	F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	22/03/2019	
					CT449	Generators and associated devices	F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	22/03/2019	
					CT981	Single use	F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	22/03/2019	
					CT983	Surgical invasive	F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	22/03/2019	
8652	2020	Death	37265	Dual-chamber implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E0611	Heart Failure/Congestive Heart Failure	A24	No Apparent Device Problem	19/01/2020	19/01/2020
							F02	Death	A24	No Apparent Device Problem	19/01/2020	19/01/2020
							F21	Unexpected Deterioration	A24	No Apparent Device Problem	19/01/2020	19/01/2020
8653	2019	All other reportable incidents	47267	Single-chamber implantable pacemaker, rate-responsive	CT141	Electrodes/Leads and associated devices	F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	12/04/2019	
					CT145	Cardiovascular devices	F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	12/04/2019	
					CT1684	Cardiac pacemakers and associated devices	F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	12/04/2019	
					CT1685	Cardiac pacemakers	F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	12/04/2019	
					CT233	Surgical	F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	12/04/2019	
					CT2406	Implantable	F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	12/04/2019	
					CT2407	Active implantable	F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	12/04/2019	
					CT257	Implantable cardiac pacemakers	F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	12/04/2019	
					CT2727	Active implant devices/ancillaries	F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	12/04/2019	
					CT2750	Implantable cardiac electrical stimulators and associated devices	F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	12/04/2019	
					CT2866	Implantable cardiac electrical stimulators	F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	12/04/2019	
					CT2985	Active	F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	12/04/2019	
					CT3001	Implantable metal	F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	12/04/2019	
					CT3002	Implantable synthetic polymer	F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	12/04/2019	
					CT3006	Implantable inorganic material	F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	12/04/2019	
					CT314	Electrically powered	F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	12/04/2019	
					CT321	Long-term surgical invasive	F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	12/04/2019	
					CT449	Generators and associated devices	F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	12/04/2019	
					CT981	Single use	F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	12/04/2019	
					CT983	Surgical invasive	F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	12/04/2019	
8654	2019	All other reportable incidents	47265	Dual-chamber implantable pacemaker, rate-responsive	CT141	Electrodes/Leads and associated devices	F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	11/03/2019	
					CT145	Cardiovascular devices	F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	11/03/2019	
					CT1684	Cardiac pacemakers and associated devices	F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	11/03/2019	
					CT1685	Cardiac pacemakers	F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	11/03/2019	
					CT233	Surgical	F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	11/03/2019	
					CT2406	Implantable	F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	11/03/2019	
					CT2407	Active implantable	F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	11/03/2019	
					CT257	Implantable cardiac pacemakers	F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	11/03/2019	
					CT2727	Active implant devices/ancillaries	F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	11/03/2019	
					CT2750	Implantable cardiac electrical stimulators and associated devices	F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	11/03/2019	

					CT2866	Implantable cardiac electrical stimulators	F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	11/03/2019		
					CT2985	Active	F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	11/03/2019		
					CT3001	Implantable metal	F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	11/03/2019		
					CT3002	Implantable synthetic polymer	F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	11/03/2019		
					CT3006	Implantable inorganic material	F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	11/03/2019		
					CT314	Electrically powered	F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	11/03/2019		
					CT321	Long-term surgical invasive	F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	11/03/2019		
					CT449	Generators and associated devices	F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	11/03/2019		
					CT981	Single use	F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	11/03/2019		
					CT983	Surgical invasive	F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	11/03/2019		
8655	2019	All other reportable incidents	35852	Single-chamber implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A070909	Oversensing			
8656	2019	All other reportable incidents	47267	Single-chamber implantable pacemaker, rate-responsive	CT141	Electrodes/Leads and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A071102	Failure to Interrogate	07/03/2019		
					CT145	Cardiovascular devices	E2403	No Clinical Signs, Symptoms or Conditions	A071102	Failure to Interrogate	07/03/2019		
					CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A071102	Failure to Interrogate	07/03/2019		
					CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A071102	Failure to Interrogate	07/03/2019		
					CT233	Surgical	E2403	No Clinical Signs, Symptoms or Conditions	A071102	Failure to Interrogate	07/03/2019		
					CT2406	Implantable	E2403	No Clinical Signs, Symptoms or Conditions	A071102	Failure to Interrogate	07/03/2019		
					CT2407	Active implantable	E2403	No Clinical Signs, Symptoms or Conditions	A071102	Failure to Interrogate	07/03/2019		
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A071102	Failure to Interrogate	07/03/2019		
					CT2727	Active implant devices/ancillaries	E2403	No Clinical Signs, Symptoms or Conditions	A071102	Failure to Interrogate	07/03/2019		
					CT2750	Implantable cardiac electrical stimulators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A071102	Failure to Interrogate	07/03/2019		
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A071102	Failure to Interrogate	07/03/2019		
					CT2985	Active	E2403	No Clinical Signs, Symptoms or Conditions	A071102	Failure to Interrogate	07/03/2019		
					CT3001	Implantable metal	E2403	No Clinical Signs, Symptoms or Conditions	A071102	Failure to Interrogate	07/03/2019		
					CT3002	Implantable synthetic polymer	E2403	No Clinical Signs, Symptoms or Conditions	A071102	Failure to Interrogate	07/03/2019		
					CT3006	Implantable inorganic material	E2403	No Clinical Signs, Symptoms or Conditions	A071102	Failure to Interrogate	07/03/2019		
					CT314	Electrically powered	E2403	No Clinical Signs, Symptoms or Conditions	A071102	Failure to Interrogate	07/03/2019		
					CT321	Long-term surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A071102	Failure to Interrogate	07/03/2019		
					CT449	Generators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A071102	Failure to Interrogate	07/03/2019		
					CT981	Single use	E2403	No Clinical Signs, Symptoms or Conditions	A071102	Failure to Interrogate	07/03/2019		
					CT983	Surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A071102	Failure to Interrogate	07/03/2019		
8657	2020	All other reportable incidents	37265	Dual-chamber implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A072201	High Impedance			
							F26	No Health Consequences or Impact	A072201	High Impedance			
8658	2020	All other reportable incidents	36241	Pacemaker/defibrillator lead	CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A072201	High Impedance			
							F26	No Health Consequences or Impact	A072201	High Impedance			
8659	2020	All other reportable incidents	47270	Cardiac resynchronization therapy implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A070909	Oversensing			
									A071302	Inappropriate/Inadequate Shock/Stimulation			
							F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A070909	Oversensing			
									A071302	Inappropriate/Inadequate Shock/Stimulation			
							F26	No Health Consequences or Impact	A070909	Oversensing			
									A071302	Inappropriate/Inadequate Shock/Stimulation			
8660	2020	All other reportable incidents	47265	Dual-chamber implantable pacemaker, rate-responsive	CT141	Electrodes/Leads and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	17/01/2020	18/02/2020	
					CT145	Cardiovascular devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	17/01/2020	18/02/2020	
					CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	17/01/2020	18/02/2020	
					CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	17/01/2020	18/02/2020	
					CT233	Surgical	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	17/01/2020	18/02/2020	
					CT2406	Implantable	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	17/01/2020	18/02/2020	
					CT2407	Active implantable	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	17/01/2020	18/02/2020	
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	17/01/2020	18/02/2020	
					CT2727	Active implant devices/ancillaries	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	17/01/2020	18/02/2020	
					CT2750	Implantable cardiac electrical stimulators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	17/01/2020	18/02/2020	
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	17/01/2020	18/02/2020	
					CT2985	Active	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	17/01/2020	18/02/2020	
					CT3001	Implantable metal	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	17/01/2020	18/02/2020	
					CT3002	Implantable synthetic polymer	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	17/01/2020	18/02/2020	
					CT3006	Implantable inorganic material	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	17/01/2020	18/02/2020	
					CT314	Electrically powered	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	17/01/2020	18/02/2020	
					CT321	Long-term surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	17/01/2020	18/02/2020	
					CT449	Generators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	17/01/2020	18/02/2020	
					CT981	Single use	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	17/01/2020	18/02/2020	
					CT983	Surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	17/01/2020	18/02/2020	
8661	2020	All other reportable incidents	47265	Dual-chamber implantable pacemaker, rate-responsive	CT141	Electrodes/Leads and associated devices	F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	22/01/2020	22/01/2020	
					CT145	Cardiovascular devices	F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	22/01/2020	22/01/2020	
					CT1684	Cardiac pacemakers and associated devices	F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	22/01/2020	22/01/2020	
					CT1685	Cardiac pacemakers	F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	22/01/2020	22/01/2020	
					CT233	Surgical	F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	22/01/2020	22/01/2020	
					CT2406	Implantable	F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	22/01/2020	22/01/2020	
					CT2407	Active implantable	F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	22/01/2020	22/01/2020	
					CT257	Implantable cardiac pacemakers	F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	22/01/2020	22/01/2020	
					CT2727	Active implant devices/ancillaries	F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	22/01/2020	22/01/2020	
					CT2750	Implantable cardiac electrical stimulators and associated devices	F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	22/01/2020	22/01/2020	
					CT2866	Implantable cardiac electrical stimulators	F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	22/01/2020	22/01/2020	
					CT2985	Active	F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	22/01/2020	22/01/2020	
					CT3001	Implantable metal	F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	22/01/2020	22/01/2020	
					CT3002	Implantable synthetic polymer	F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	22/01/2020	22/01/2020	
					CT3006	Implantable inorganic material	F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	22/01/2020	22/01/2020	
					CT314	Electrically powered	F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	22/01/2020	22/01/2020	
					CT321	Long-term surgical invasive	F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	22/01/2020	22/01/2020	
					CT449	Generators and associated devices	F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	22/01/2020	22/01/2020	
					CT981	Single use	F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	22/01/2020	22/01/2020	
					CT983	Surgical invasive	F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	22/01/2020	22/01/2020	
8662	2019	All other reportable incidents	35223	Endocardial/interventricular septal pacing lead	CT1684	Cardiac pacemakers and associated devices	F1905	Device Revision or Replacement	A0404	Crack			
8663	2019	All other reportable incidents	36241	Pacemaker/defibrillator lead	CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A072202	Low Impedance	11/03/2019		
8664	2019	All other reportable incidents	47265	Dual-chamber implantable pacemaker, rate-responsive	CT141	Electrodes/Leads and associated devices	F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery			
					CT145	Cardiovascular devices	F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery			

					CT1684	Cardiac pacemakers and associated devices	F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery		
					CT1685	Cardiac pacemakers	F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery		
					CT233	Surgical	F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery		
					CT2406	Implantable	F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery		
					CT2407	Active implantable	F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery		
					CT257	Implantable cardiac pacemakers	F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery		
					CT2727	Active implant devices/ancillaries	F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery		
					CT2750	Implantable cardiac electrical stimulators and associated devices	F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery		
					CT2866	Implantable cardiac electrical stimulators	F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery		
					CT2985	Active	F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery		
					CT3001	Implantable metal	F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery		
					CT3002	Implantable synthetic polymer	F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery		
					CT3006	Implantable inorganic material	F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery		
					CT314	Electrically powered	F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery		
					CT321	Long-term surgical invasive	F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery		
					CT449	Generators and associated devices	F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery		
					CT981	Single use	F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery		
					CT983	Surgical invasive	F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery		
8665	2019	All other reportable incidents	36241	Pacemaker/defibrillator lead	CT1684	Cardiac pacemakers and associated devices	F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A070909	Oversensing		
							F14	Prolonged Episode of Care	A070909	Oversensing		
8666	2019	All other reportable incidents	36241	Pacemaker/defibrillator lead	CT1684	Cardiac pacemakers and associated devices	F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A070909	Oversensing		
8667	2019	All other reportable incidents	35852	Single-chamber implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A071302	Inappropriate/Inadequate Shock/Stimulation		
							F1905	Device Revision or Replacement	A071302	Inappropriate/Inadequate Shock/Stimulation		
8668	2020	All other reportable incidents	47265	Dual-chamber implantable pacemaker, rate-responsive	CT141	Electrodes/Leads and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A072202	Low Impedance	23/01/2020	23/01/2020
									A180103	Device Contamination with Body Fluid	23/01/2020	23/01/2020
							F1901	Additional Surgery	A072202	Low Impedance	23/01/2020	23/01/2020
									A180103	Device Contamination with Body Fluid	23/01/2020	23/01/2020
							F1905	Device Revision or Replacement	A072202	Low Impedance	23/01/2020	23/01/2020
									A180103	Device Contamination with Body Fluid	23/01/2020	23/01/2020
					CT145	Cardiovascular devices	E2403	No Clinical Signs, Symptoms or Conditions	A072202	Low Impedance	23/01/2020	23/01/2020
									A180103	Device Contamination with Body Fluid	23/01/2020	23/01/2020
							F1901	Additional Surgery	A072202	Low Impedance	23/01/2020	23/01/2020
									A180103	Device Contamination with Body Fluid	23/01/2020	23/01/2020
							F1905	Device Revision or Replacement	A072202	Low Impedance	23/01/2020	23/01/2020
									A180103	Device Contamination with Body Fluid	23/01/2020	23/01/2020
					CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A072202	Low Impedance	23/01/2020	23/01/2020
									A180103	Device Contamination with Body Fluid	23/01/2020	23/01/2020
							F1901	Additional Surgery	A072202	Low Impedance	23/01/2020	23/01/2020
									A180103	Device Contamination with Body Fluid	23/01/2020	23/01/2020
							F1905	Device Revision or Replacement	A072202	Low Impedance	23/01/2020	23/01/2020
									A180103	Device Contamination with Body Fluid	23/01/2020	23/01/2020
					CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A072202	Low Impedance	23/01/2020	23/01/2020
									A180103	Device Contamination with Body Fluid	23/01/2020	23/01/2020
							F1901	Additional Surgery	A072202	Low Impedance	23/01/2020	23/01/2020
									A180103	Device Contamination with Body Fluid	23/01/2020	23/01/2020
							F1905	Device Revision or Replacement	A072202	Low Impedance	23/01/2020	23/01/2020
									A180103	Device Contamination with Body Fluid	23/01/2020	23/01/2020
					CT233	Surgical	E2403	No Clinical Signs, Symptoms or Conditions	A072202	Low Impedance	23/01/2020	23/01/2020
									A180103	Device Contamination with Body Fluid	23/01/2020	23/01/2020
							F1901	Additional Surgery	A072202	Low Impedance	23/01/2020	23/01/2020
									A180103	Device Contamination with Body Fluid	23/01/2020	23/01/2020
							F1905	Device Revision or Replacement	A072202	Low Impedance	23/01/2020	23/01/2020
									A180103	Device Contamination with Body Fluid	23/01/2020	23/01/2020
					CT2406	Implantable	E2403	No Clinical Signs, Symptoms or Conditions	A072202	Low Impedance	23/01/2020	23/01/2020
									A180103	Device Contamination with Body Fluid	23/01/2020	23/01/2020
							F1901	Additional Surgery	A072202	Low Impedance	23/01/2020	23/01/2020
									A180103	Device Contamination with Body Fluid	23/01/2020	23/01/2020
							F1905	Device Revision or Replacement	A072202	Low Impedance	23/01/2020	23/01/2020
									A180103	Device Contamination with Body Fluid	23/01/2020	23/01/2020
					CT2407	Active implantable	E2403	No Clinical Signs, Symptoms or Conditions	A072202	Low Impedance	23/01/2020	23/01/2020
									A180103	Device Contamination with Body Fluid	23/01/2020	23/01/2020
							F1901	Additional Surgery	A072202	Low Impedance	23/01/2020	23/01/2020
									A180103	Device Contamination with Body Fluid	23/01/2020	23/01/2020
							F1905	Device Revision or Replacement	A072202	Low Impedance	23/01/2020	23/01/2020
									A180103	Device Contamination with Body Fluid	23/01/2020	23/01/2020
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A072202	Low Impedance	23/01/2020	23/01/2020
									A180103	Device Contamination with Body Fluid	23/01/2020	23/01/2020
							F1901	Additional Surgery	A072202	Low Impedance	23/01/2020	23/01/2020
									A180103	Device Contamination with Body Fluid	23/01/2020	23/01/2020
							F1905	Device Revision or Replacement	A072202	Low Impedance	23/01/2020	23/01/2020
									A180103	Device Contamination with Body Fluid	23/01/2020	23/01/2020
					CT2727	Active implant devices/ancillaries	E2403	No Clinical Signs, Symptoms or Conditions	A072202	Low Impedance	23/01/2020	23/01/2020
									A180103	Device Contamination with Body Fluid	23/01/2020	23/01/2020
							F1901	Additional Surgery	A072202	Low Impedance	23/01/2020	23/01/2020
									A180103	Device Contamination with Body Fluid	23/01/2020	23/01/2020
							F1905	Device Revision or Replacement	A072202	Low Impedance	23/01/2020	23/01/2020
									A180103	Device Contamination with Body Fluid	23/01/2020	23/01/2020
					CT2750	Implantable cardiac electrical stimulators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A072202	Low Impedance	23/01/2020	23/01/2020
									A180103	Device Contamination with Body Fluid	23/01/2020	23/01/2020
							F1901	Additional Surgery	A072202	Low Impedance	23/01/2020	23/01/2020
									A180103	Device Contamination with Body Fluid	23/01/2020	23/01/2020
							F1905	Device Revision or Replacement	A072202	Low Impedance	23/01/2020	23/01/2020
									A180103	Device Contamination with Body Fluid	23/01/2020	23/01/2020
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A072202	Low Impedance	23/01/2020	23/01/2020
									A180103	Device Contamination with Body Fluid	23/01/2020	23/01/2020
							F1901	Additional Surgery	A072202	Low Impedance	23/01/2020	23/01/2020
									A180103	Device Contamination with Body Fluid	23/01/2020	23/01/2020

							F1905	Device Revision or Replacement	A072202	Low Impedance	23/01/2020	23/01/2020
									A180103	Device Contamination with Body Fluid	23/01/2020	23/01/2020
					CT2985	Active	E2403	No Clinical Signs, Symptoms or Conditions	A072202	Low Impedance	23/01/2020	23/01/2020
									A180103	Device Contamination with Body Fluid	23/01/2020	23/01/2020
							F1901	Additional Surgery	A072202	Low Impedance	23/01/2020	23/01/2020
									A180103	Device Contamination with Body Fluid	23/01/2020	23/01/2020
							F1905	Device Revision or Replacement	A072202	Low Impedance	23/01/2020	23/01/2020
									A180103	Device Contamination with Body Fluid	23/01/2020	23/01/2020
					CT3001	Implantable metal	E2403	No Clinical Signs, Symptoms or Conditions	A072202	Low Impedance	23/01/2020	23/01/2020
									A180103	Device Contamination with Body Fluid	23/01/2020	23/01/2020
							F1901	Additional Surgery	A072202	Low Impedance	23/01/2020	23/01/2020
									A180103	Device Contamination with Body Fluid	23/01/2020	23/01/2020
							F1905	Device Revision or Replacement	A072202	Low Impedance	23/01/2020	23/01/2020
									A180103	Device Contamination with Body Fluid	23/01/2020	23/01/2020
					CT3002	Implantable synthetic polymer	E2403	No Clinical Signs, Symptoms or Conditions	A072202	Low Impedance	23/01/2020	23/01/2020
									A180103	Device Contamination with Body Fluid	23/01/2020	23/01/2020
							F1901	Additional Surgery	A072202	Low Impedance	23/01/2020	23/01/2020
									A180103	Device Contamination with Body Fluid	23/01/2020	23/01/2020
							F1905	Device Revision or Replacement	A072202	Low Impedance	23/01/2020	23/01/2020
									A180103	Device Contamination with Body Fluid	23/01/2020	23/01/2020
					CT3006	Implantable inorganic material	E2403	No Clinical Signs, Symptoms or Conditions	A072202	Low Impedance	23/01/2020	23/01/2020
									A180103	Device Contamination with Body Fluid	23/01/2020	23/01/2020
							F1901	Additional Surgery	A072202	Low Impedance	23/01/2020	23/01/2020
									A180103	Device Contamination with Body Fluid	23/01/2020	23/01/2020
							F1905	Device Revision or Replacement	A072202	Low Impedance	23/01/2020	23/01/2020
									A180103	Device Contamination with Body Fluid	23/01/2020	23/01/2020
					CT314	Electrically powered	E2403	No Clinical Signs, Symptoms or Conditions	A072202	Low Impedance	23/01/2020	23/01/2020
									A180103	Device Contamination with Body Fluid	23/01/2020	23/01/2020
							F1901	Additional Surgery	A072202	Low Impedance	23/01/2020	23/01/2020
									A180103	Device Contamination with Body Fluid	23/01/2020	23/01/2020
							F1905	Device Revision or Replacement	A072202	Low Impedance	23/01/2020	23/01/2020
									A180103	Device Contamination with Body Fluid	23/01/2020	23/01/2020
					CT321	Long-term surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A072202	Low Impedance	23/01/2020	23/01/2020
									A180103	Device Contamination with Body Fluid	23/01/2020	23/01/2020
							F1901	Additional Surgery	A072202	Low Impedance	23/01/2020	23/01/2020
									A180103	Device Contamination with Body Fluid	23/01/2020	23/01/2020
							F1905	Device Revision or Replacement	A072202	Low Impedance	23/01/2020	23/01/2020
									A180103	Device Contamination with Body Fluid	23/01/2020	23/01/2020
					CT449	Generators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A072202	Low Impedance	23/01/2020	23/01/2020
									A180103	Device Contamination with Body Fluid	23/01/2020	23/01/2020
							F1901	Additional Surgery	A072202	Low Impedance	23/01/2020	23/01/2020
									A180103	Device Contamination with Body Fluid	23/01/2020	23/01/2020
							F1905	Device Revision or Replacement	A072202	Low Impedance	23/01/2020	23/01/2020
									A180103	Device Contamination with Body Fluid	23/01/2020	23/01/2020
					CT981	Single use	E2403	No Clinical Signs, Symptoms or Conditions	A072202	Low Impedance	23/01/2020	23/01/2020
									A180103	Device Contamination with Body Fluid	23/01/2020	23/01/2020
							F1901	Additional Surgery	A072202	Low Impedance	23/01/2020	23/01/2020
									A180103	Device Contamination with Body Fluid	23/01/2020	23/01/2020
							F1905	Device Revision or Replacement	A072202	Low Impedance	23/01/2020	23/01/2020
									A180103	Device Contamination with Body Fluid	23/01/2020	23/01/2020
					CT983	Surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A072202	Low Impedance	23/01/2020	23/01/2020
									A180103	Device Contamination with Body Fluid	23/01/2020	23/01/2020
							F1901	Additional Surgery	A072202	Low Impedance	23/01/2020	23/01/2020
									A180103	Device Contamination with Body Fluid	23/01/2020	23/01/2020
							F1905	Device Revision or Replacement	A072202	Low Impedance	23/01/2020	23/01/2020
									A180103	Device Contamination with Body Fluid	23/01/2020	23/01/2020
8669	2020	All other reportable incidents	36241	Pacemaker/defibrillator lead	CT1684	Cardiac pacemakers and associated devices	E2114	Perforation	A0712	Pacing Problem		
							F12	Serious Injury/ Illness/ Impairment	A0712	Pacing Problem		
							F19	Surgical Intervention	A0712	Pacing Problem		
							F1905	Device Revision or Replacement	A0712	Pacing Problem		
8670	2020	All other reportable incidents	47265	Dual-chamber implantable pacemaker, rate-responsive	CT141	Electrodes/Leads and associated devices	F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	17/02/2020	17/02/2020
					CT145	Cardiovascular devices	F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	17/02/2020	17/02/2020
					CT1684	Cardiac pacemakers and associated devices	F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	17/02/2020	17/02/2020
					CT1685	Cardiac pacemakers	F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	17/02/2020	17/02/2020
					CT233	Surgical	F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	17/02/2020	17/02/2020
					CT2406	Implantable	F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	17/02/2020	17/02/2020
					CT2407	Active implantable	F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	17/02/2020	17/02/2020
					CT257	Implantable cardiac pacemakers	F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	17/02/2020	17/02/2020
					CT2727	Active implant devices/ancillaries	F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	17/02/2020	17/02/2020
					CT2750	Implantable cardiac electrical stimulators and associated devices	F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	17/02/2020	17/02/2020
					CT2866	Implantable cardiac electrical stimulators	F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	17/02/2020	17/02/2020
					CT2985	Active	F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	17/02/2020	17/02/2020
					CT3001	Implantable metal	F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	17/02/2020	17/02/2020
					CT3002	Implantable synthetic polymer	F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	17/02/2020	17/02/2020
					CT3006	Implantable inorganic material	F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	17/02/2020	17/02/2020
					CT314	Electrically powered	F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	17/02/2020	17/02/2020
					CT321	Long-term surgical invasive	F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	17/02/2020	17/02/2020
					CT449	Generators and associated devices	F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	17/02/2020	17/02/2020
					CT981	Single use	F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	17/02/2020	17/02/2020
					CT983	Surgical invasive	F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	17/02/2020	17/02/2020
8671	2020	All other reportable incidents	47265	Dual-chamber implantable pacemaker, rate-responsive	CT141	Electrodes/Leads and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	20/01/2020	20/01/2020
							F19	Surgical Intervention	A070504	Premature Discharge of Battery	20/01/2020	20/01/2020
					CT145	Cardiovascular devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	20/01/2020	20/01/2020
							F19	Surgical Intervention	A070504	Premature Discharge of Battery	20/01/2020	20/01/2020
					CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	20/01/2020	20/01/2020
							F19	Surgical Intervention	A070504	Premature Discharge of Battery	20/01/2020	20/01/2020
					CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	20/01/2020	20/01/2020

							F19	Surgical Intervention	A070504	Premature Discharge of Battery	20/01/2020	20/01/2020
					CT233	Surgical	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	20/01/2020	20/01/2020
							F19	Surgical Intervention	A070504	Premature Discharge of Battery	20/01/2020	20/01/2020
					CT2406	Implantable	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	20/01/2020	20/01/2020
							F19	Surgical Intervention	A070504	Premature Discharge of Battery	20/01/2020	20/01/2020
					CT2407	Active implantable	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	20/01/2020	20/01/2020
							F19	Surgical Intervention	A070504	Premature Discharge of Battery	20/01/2020	20/01/2020
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	20/01/2020	20/01/2020
							F19	Surgical Intervention	A070504	Premature Discharge of Battery	20/01/2020	20/01/2020
					CT2727	Active implant devices/ancillaries	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	20/01/2020	20/01/2020
							F19	Surgical Intervention	A070504	Premature Discharge of Battery	20/01/2020	20/01/2020
					CT2750	Implantable cardiac electrical stimulators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	20/01/2020	20/01/2020
							F19	Surgical Intervention	A070504	Premature Discharge of Battery	20/01/2020	20/01/2020
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	20/01/2020	20/01/2020
							F19	Surgical Intervention	A070504	Premature Discharge of Battery	20/01/2020	20/01/2020
					CT2985	Active	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	20/01/2020	20/01/2020
							F19	Surgical Intervention	A070504	Premature Discharge of Battery	20/01/2020	20/01/2020
					CT3001	Implantable metal	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	20/01/2020	20/01/2020
							F19	Surgical Intervention	A070504	Premature Discharge of Battery	20/01/2020	20/01/2020
					CT3002	Implantable synthetic polymer	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	20/01/2020	20/01/2020
							F19	Surgical Intervention	A070504	Premature Discharge of Battery	20/01/2020	20/01/2020
					CT3006	Implantable inorganic material	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	20/01/2020	20/01/2020
							F19	Surgical Intervention	A070504	Premature Discharge of Battery	20/01/2020	20/01/2020
					CT314	Electrically powered	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	20/01/2020	20/01/2020
							F19	Surgical Intervention	A070504	Premature Discharge of Battery	20/01/2020	20/01/2020
					CT321	Long-term surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	20/01/2020	20/01/2020
							F19	Surgical Intervention	A070504	Premature Discharge of Battery	20/01/2020	20/01/2020
					CT449	Generators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	20/01/2020	20/01/2020
							F19	Surgical Intervention	A070504	Premature Discharge of Battery	20/01/2020	20/01/2020
					CT981	Single use	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	20/01/2020	20/01/2020
							F19	Surgical Intervention	A070504	Premature Discharge of Battery	20/01/2020	20/01/2020
					CT983	Surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	20/01/2020	20/01/2020
							F19	Surgical Intervention	A070504	Premature Discharge of Battery	20/01/2020	20/01/2020
8672	2020	All other reportable incidents	37265	Dual-chamber implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	07/02/2020	07/02/2020
							F19	Surgical Intervention	A070504	Premature Discharge of Battery	07/02/2020	07/02/2020
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	07/02/2020	07/02/2020
8673	2020	All other reportable incidents	35223	Endocardial/interventricular septal pacing lead	CT1684	Cardiac pacemakers and associated devices	E2401	Insufficient Information	A051201	Device Dislodged or Dislocated		
							F1904	Device Repositioning	A051201	Device Dislodged or Dislocated		
							F1905	Device Revision or Replacement	A051201	Device Dislodged or Dislocated		
8674	2020	All other reportable incidents	35223	Endocardial/interventricular septal pacing lead	CT1684	Cardiac pacemakers and associated devices	E2401	Insufficient Information	A051201	Device Dislodged or Dislocated		
							F1904	Device Repositioning	A051201	Device Dislodged or Dislocated		
							F1905	Device Revision or Replacement	A051201	Device Dislodged or Dislocated		
8675	2020	All other reportable incidents	35223	Endocardial/interventricular septal pacing lead	CT1684	Cardiac pacemakers and associated devices	E2401	Insufficient Information	A051201	Device Dislodged or Dislocated		
							F1904	Device Repositioning	A051201	Device Dislodged or Dislocated		
							F1905	Device Revision or Replacement	A051201	Device Dislodged or Dislocated		
8676	2020	All other reportable incidents	35223	Endocardial/interventricular septal pacing lead	CT1684	Cardiac pacemakers and associated devices	E2401	Insufficient Information	A051201	Device Dislodged or Dislocated		
							F1904	Device Repositioning	A051201	Device Dislodged or Dislocated		
							F1905	Device Revision or Replacement	A051201	Device Dislodged or Dislocated		
8677	2020	All other reportable incidents	35223	Endocardial/interventricular septal pacing lead	CT1684	Cardiac pacemakers and associated devices	E2401	Insufficient Information	A051201	Device Dislodged or Dislocated		
							E2403	No Clinical Signs, Symptoms or Conditions	A051201	Device Dislodged or Dislocated		
							F1904	Device Repositioning	A051201	Device Dislodged or Dislocated		
8678	2020	All other reportable incidents	35223	Endocardial/interventricular septal pacing lead	CT1684	Cardiac pacemakers and associated devices	E2401	Insufficient Information	A051201	Device Dislodged or Dislocated		
							F1904	Device Repositioning	A051201	Device Dislodged or Dislocated		
							F1905	Device Revision or Replacement	A051201	Device Dislodged or Dislocated		
8679	2020	All other reportable incidents	35223	Endocardial/interventricular septal pacing lead	CT1684	Cardiac pacemakers and associated devices	E2401	Insufficient Information	A051201	Device Dislodged or Dislocated	03/05/2019	
							F1903	Device Explantation	A051201	Device Dislodged or Dislocated	03/05/2019	
							F1905	Device Revision or Replacement	A051201	Device Dislodged or Dislocated	03/05/2019	
8680	2020	All other reportable incidents	35223	Endocardial/interventricular septal pacing lead	CT1684	Cardiac pacemakers and associated devices	E2401	Insufficient Information	A051201	Device Dislodged or Dislocated		
							F1904	Device Repositioning	A051201	Device Dislodged or Dislocated		
							F1905	Device Revision or Replacement	A051201	Device Dislodged or Dislocated		
8681	2019	All other reportable incidents	47265	Dual-chamber implantable pacemaker, rate-responsive	CT141	Electrodes/Leads and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	20/03/2019	
					CT145	Cardiovascular devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	20/03/2019	
					CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	20/03/2019	
					CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	20/03/2019	
					CT233	Surgical	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	20/03/2019	
					CT2406	Implantable	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	20/03/2019	
					CT2407	Active implantable	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	20/03/2019	
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	20/03/2019	
					CT2727	Active implant devices/ancillaries	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	20/03/2019	
					CT2750	Implantable cardiac electrical stimulators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	20/03/2019	
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	20/03/2019	
					CT2985	Active	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	20/03/2019	
					CT3001	Implantable metal	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	20/03/2019	
					CT3002	Implantable synthetic polymer	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	20/03/2019	
					CT3006	Implantable inorganic material	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	20/03/2019	
					CT314	Electrically powered	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	20/03/2019	
					CT321	Long-term surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	20/03/2019	
					CT449	Generators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	20/03/2019	
					CT981	Single use	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	20/03/2019	
					CT983	Surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	20/03/2019	
8682	2019	All other reportable incidents	17882	Physiologic-monitoring defibrillation system	CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A0707	Failure to Discharge		
8683	2019	All other reportable incidents	17882	Physiologic-monitoring defibrillation system	CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A0907	No Device Output		
8684	2019	All other reportable incidents	47265	Dual-chamber implantable pacemaker, rate-responsive	CT141	Electrodes/Leads and associated devices	F1905	Device Revision or Replacement	A0721	Circuit Failure	11/03/2019	
					CT145	Cardiovascular devices	F1905	Device Revision or Replacement	A0721	Circuit Failure	11/03/2019	
					CT1684	Cardiac pacemakers and associated devices	F1905	Device Revision or Replacement	A0721	Circuit Failure	11/03/2019	
					CT1685	Cardiac pacemakers	F1905	Device Revision or Replacement	A0721	Circuit Failure	11/03/2019	
					CT233	Surgical	F1905	Device Revision or Replacement	A0721	Circuit Failure	11/03/2019	

					CT2406	Implantable	F1905	Device Revision or Replacement	A0721	Circuit Failure	11/03/2019		
					CT2407	Active implantable	F1905	Device Revision or Replacement	A0721	Circuit Failure	11/03/2019		
					CT257	Implantable cardiac pacemakers	F1905	Device Revision or Replacement	A0721	Circuit Failure	11/03/2019		
					CT2727	Active implant devices/ancillaries	F1905	Device Revision or Replacement	A0721	Circuit Failure	11/03/2019		
					CT2750	Implantable cardiac electrical stimulators and associated devices	F1905	Device Revision or Replacement	A0721	Circuit Failure	11/03/2019		
					CT2866	Implantable cardiac electrical stimulators	F1905	Device Revision or Replacement	A0721	Circuit Failure	11/03/2019		
					CT2985	Active	F1905	Device Revision or Replacement	A0721	Circuit Failure	11/03/2019		
					CT3001	Implantable metal	F1905	Device Revision or Replacement	A0721	Circuit Failure	11/03/2019		
					CT3002	Implantable synthetic polymer	F1905	Device Revision or Replacement	A0721	Circuit Failure	11/03/2019		
					CT3006	Implantable inorganic material	F1905	Device Revision or Replacement	A0721	Circuit Failure	11/03/2019		
					CT314	Electrically powered	F1905	Device Revision or Replacement	A0721	Circuit Failure	11/03/2019		
					CT321	Long-term surgical invasive	F1905	Device Revision or Replacement	A0721	Circuit Failure	11/03/2019		
					CT449	Generators and associated devices	F1905	Device Revision or Replacement	A0721	Circuit Failure	11/03/2019		
					CT981	Single use	F1905	Device Revision or Replacement	A0721	Circuit Failure	11/03/2019		
					CT983	Surgical invasive	F1905	Device Revision or Replacement	A0721	Circuit Failure	11/03/2019		
8685	2019	All other reportable incidents	47270	Cardiac resynchronization therapy implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem	02/04/2019		
8686	2019	All other reportable incidents	37265	Dual-chamber implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	F1905	Device Revision or Replacement	A0706	Charging Problem	29/03/2019		
8687	2019	All other reportable incidents	35852	Single-chamber implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	05/04/2019		
8688	2019	All other reportable incidents	36241	Pacemaker/defibrillator lead	CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A0510	Retraction Problem	07/03/2019		
8689	2019	All other reportable incidents	37265	Dual-chamber implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A070909	Oversensing			
8690	2020	All other reportable incidents	47263	Cardiac resynchronization therapy implantable pacemaker	CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A072201	High Impedance	29/01/2020	29/01/2020	
							F1901	Additional Surgery	A072201	High Impedance	29/01/2020	29/01/2020	
							F1905	Device Revision or Replacement	A072201	High Impedance	29/01/2020	29/01/2020	
					CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A072201	High Impedance	29/01/2020	29/01/2020	
							F1901	Additional Surgery	A072201	High Impedance	29/01/2020	29/01/2020	
							F1905	Device Revision or Replacement	A072201	High Impedance	29/01/2020	29/01/2020	
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A072201	High Impedance	29/01/2020	29/01/2020	
							F1901	Additional Surgery	A072201	High Impedance	29/01/2020	29/01/2020	
							F1905	Device Revision or Replacement	A072201	High Impedance	29/01/2020	29/01/2020	
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A072201	High Impedance	29/01/2020	29/01/2020	
							F1901	Additional Surgery	A072201	High Impedance	29/01/2020	29/01/2020	
							F1905	Device Revision or Replacement	A072201	High Impedance	29/01/2020	29/01/2020	
8691	2019	All other reportable incidents	36241	Pacemaker/defibrillator lead	CT1684	Cardiac pacemakers and associated devices	F05	Delay to Treatment/ Therapy	A0510	Retraction Problem			
8692	2020	All other reportable incidents	37265	Dual-chamber implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	03/02/2020	03/02/2020	
							F1901	Additional Surgery	A070504	Premature Discharge of Battery	03/02/2020	03/02/2020	
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	03/02/2020	03/02/2020	
8693	2019	All other reportable incidents	36241	Pacemaker/defibrillator lead	CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A0510	Retraction Problem			
8694	2019	All other reportable incidents	36241	Pacemaker/defibrillator lead	CT1684	Cardiac pacemakers and associated devices	F1905	Device Revision or Replacement	A0510	Retraction Problem	14/03/2019		
8695	2019	All other reportable incidents	36241	Pacemaker/defibrillator lead	CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A051201	Device Dislodged or Dislocated			
8696	2019	All other reportable incidents	36241	Pacemaker/defibrillator lead	CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A	Annex A			
8697	2020	All other reportable incidents	47265	Dual-chamber implantable pacemaker, rate-responsive	CT141	Electrodes/Leads and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A0722	Impedance Problem			
							F26	No Health Consequences or Impact	A0722	Impedance Problem			
					CT145	Cardiovascular devices	E2403	No Clinical Signs, Symptoms or Conditions	A0722	Impedance Problem			
							F26	No Health Consequences or Impact	A0722	Impedance Problem			
					CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A0722	Impedance Problem			
							F26	No Health Consequences or Impact	A0722	Impedance Problem			
					CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A0722	Impedance Problem			
							F26	No Health Consequences or Impact	A0722	Impedance Problem			
					CT233	Surgical	E2403	No Clinical Signs, Symptoms or Conditions	A0722	Impedance Problem			
							F26	No Health Consequences or Impact	A0722	Impedance Problem			
					CT2406	Implantable	E2403	No Clinical Signs, Symptoms or Conditions	A0722	Impedance Problem			
							F26	No Health Consequences or Impact	A0722	Impedance Problem			
					CT2407	Active implantable	E2403	No Clinical Signs, Symptoms or Conditions	A0722	Impedance Problem			
							F26	No Health Consequences or Impact	A0722	Impedance Problem			
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A0722	Impedance Problem			
							F26	No Health Consequences or Impact	A0722	Impedance Problem			
					CT2727	Active implant devices/ancillaries	E2403	No Clinical Signs, Symptoms or Conditions	A0722	Impedance Problem			
							F26	No Health Consequences or Impact	A0722	Impedance Problem			
					CT2750	Implantable cardiac electrical stimulators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A0722	Impedance Problem			
							F26	No Health Consequences or Impact	A0722	Impedance Problem			
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A0722	Impedance Problem			
							F26	No Health Consequences or Impact	A0722	Impedance Problem			
					CT2985	Active	E2403	No Clinical Signs, Symptoms or Conditions	A0722	Impedance Problem			
							F26	No Health Consequences or Impact	A0722	Impedance Problem			
					CT3001	Implantable metal	E2403	No Clinical Signs, Symptoms or Conditions	A0722	Impedance Problem			
							F26	No Health Consequences or Impact	A0722	Impedance Problem			
					CT3002	Implantable synthetic polymer	E2403	No Clinical Signs, Symptoms or Conditions	A0722	Impedance Problem			
							F26	No Health Consequences or Impact	A0722	Impedance Problem			
					CT3006	Implantable inorganic material	E2403	No Clinical Signs, Symptoms or Conditions	A0722	Impedance Problem			
							F26	No Health Consequences or Impact	A0722	Impedance Problem			
					CT314	Electrically powered	E2403	No Clinical Signs, Symptoms or Conditions	A0722	Impedance Problem			
							F26	No Health Consequences or Impact	A0722	Impedance Problem			
					CT321	Long-term surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A0722	Impedance Problem			
							F26	No Health Consequences or Impact	A0722	Impedance Problem			
					CT449	Generators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A0722	Impedance Problem			
							F26	No Health Consequences or Impact	A0722	Impedance Problem			
					CT981	Single use	E2403	No Clinical Signs, Symptoms or Conditions	A0722	Impedance Problem			
							F26	No Health Consequences or Impact	A0722	Impedance Problem			
					CT983	Surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A0722	Impedance Problem			
							F26	No Health Consequences or Impact	A0722	Impedance Problem			
8698	2020	All other reportable incidents	37265	Dual-chamber implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E2401	Insufficient Information	A160601	Premature Elective Replacement Indicator	10/02/2020	10/02/2020	
							E2403	No Clinical Signs, Symptoms or Conditions	A160601	Premature Elective Replacement Indicator	10/02/2020	10/02/2020	
							F1903	Device Explantation	A160601	Premature Elective Replacement Indicator	10/02/2020	10/02/2020	
8699	2020	All other reportable incidents	17882	Physiologic-monitoring defibrillation system	CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A071302	Inappropriate/Inadequate Shock/Stimulation			
							F27	Problem Identified During Non-Clinical Procedure	A071302	Inappropriate/Inadequate Shock/Stimulation			
8700	2020	All other reportable incidents	17882	Physiologic-monitoring defibrillation system	CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070802	Intermittent Loss of Power			

							F26	No Health Consequences or Impact	A070802	Intermittent Loss of Power		
8701	2020	All other reportable incidents	17882	Physiologic-monitoring defibrillation system	CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070803	Failure to Power Up		
							F27	Problem Identified During Non-Clinical Procedure	A070803	Failure to Power Up		
8702	2020	All other reportable incidents	47270	Cardiac resynchronization therapy implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	30/01/2020	30/01/2020
							F1901	Additional Surgery	A070504	Premature Discharge of Battery	30/01/2020	30/01/2020
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	30/01/2020	30/01/2020
8703	2020	All other reportable incidents	47265	Dual-chamber implantable pacemaker, rate-responsive	CT141	Electrodes/Leads and associated devices	E2311	Discomfort	A070504	Premature Discharge of Battery	28/01/2020	28/01/2020
									A071205	Pacemaker Found in Back-Up Mode	28/01/2020	28/01/2020
							F1901	Additional Surgery	A070504	Premature Discharge of Battery	28/01/2020	28/01/2020
									A071205	Pacemaker Found in Back-Up Mode	28/01/2020	28/01/2020
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	28/01/2020	28/01/2020
									A071205	Pacemaker Found in Back-Up Mode	28/01/2020	28/01/2020
					CT145	Cardiovascular devices	E2311	Discomfort	A070504	Premature Discharge of Battery	28/01/2020	28/01/2020
									A071205	Pacemaker Found in Back-Up Mode	28/01/2020	28/01/2020
							F1901	Additional Surgery	A070504	Premature Discharge of Battery	28/01/2020	28/01/2020
									A071205	Pacemaker Found in Back-Up Mode	28/01/2020	28/01/2020
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	28/01/2020	28/01/2020
									A071205	Pacemaker Found in Back-Up Mode	28/01/2020	28/01/2020
					CT1684	Cardiac pacemakers and associated devices	E2311	Discomfort	A070504	Premature Discharge of Battery	28/01/2020	28/01/2020
									A071205	Pacemaker Found in Back-Up Mode	28/01/2020	28/01/2020
							F1901	Additional Surgery	A070504	Premature Discharge of Battery	28/01/2020	28/01/2020
									A071205	Pacemaker Found in Back-Up Mode	28/01/2020	28/01/2020
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	28/01/2020	28/01/2020
									A071205	Pacemaker Found in Back-Up Mode	28/01/2020	28/01/2020
					CT1685	Cardiac pacemakers	E2311	Discomfort	A070504	Premature Discharge of Battery	28/01/2020	28/01/2020
									A071205	Pacemaker Found in Back-Up Mode	28/01/2020	28/01/2020
							F1901	Additional Surgery	A070504	Premature Discharge of Battery	28/01/2020	28/01/2020
									A071205	Pacemaker Found in Back-Up Mode	28/01/2020	28/01/2020
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	28/01/2020	28/01/2020
									A071205	Pacemaker Found in Back-Up Mode	28/01/2020	28/01/2020
					CT233	Surgical	E2311	Discomfort	A070504	Premature Discharge of Battery	28/01/2020	28/01/2020
									A071205	Pacemaker Found in Back-Up Mode	28/01/2020	28/01/2020
							F1901	Additional Surgery	A070504	Premature Discharge of Battery	28/01/2020	28/01/2020
									A071205	Pacemaker Found in Back-Up Mode	28/01/2020	28/01/2020
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	28/01/2020	28/01/2020
									A071205	Pacemaker Found in Back-Up Mode	28/01/2020	28/01/2020
					CT2406	Implantable	E2311	Discomfort	A070504	Premature Discharge of Battery	28/01/2020	28/01/2020
									A071205	Pacemaker Found in Back-Up Mode	28/01/2020	28/01/2020
							F1901	Additional Surgery	A070504	Premature Discharge of Battery	28/01/2020	28/01/2020
									A071205	Pacemaker Found in Back-Up Mode	28/01/2020	28/01/2020
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	28/01/2020	28/01/2020
									A071205	Pacemaker Found in Back-Up Mode	28/01/2020	28/01/2020
					CT2407	Active implantable	E2311	Discomfort	A070504	Premature Discharge of Battery	28/01/2020	28/01/2020
									A071205	Pacemaker Found in Back-Up Mode	28/01/2020	28/01/2020
							F1901	Additional Surgery	A070504	Premature Discharge of Battery	28/01/2020	28/01/2020
									A071205	Pacemaker Found in Back-Up Mode	28/01/2020	28/01/2020
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	28/01/2020	28/01/2020
									A071205	Pacemaker Found in Back-Up Mode	28/01/2020	28/01/2020
					CT257	Implantable cardiac pacemakers	E2311	Discomfort	A070504	Premature Discharge of Battery	28/01/2020	28/01/2020
									A071205	Pacemaker Found in Back-Up Mode	28/01/2020	28/01/2020
							F1901	Additional Surgery	A070504	Premature Discharge of Battery	28/01/2020	28/01/2020
									A071205	Pacemaker Found in Back-Up Mode	28/01/2020	28/01/2020
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	28/01/2020	28/01/2020
									A071205	Pacemaker Found in Back-Up Mode	28/01/2020	28/01/2020
					CT2727	Active implant devices/ancillaries	E2311	Discomfort	A070504	Premature Discharge of Battery	28/01/2020	28/01/2020
									A071205	Pacemaker Found in Back-Up Mode	28/01/2020	28/01/2020
							F1901	Additional Surgery	A070504	Premature Discharge of Battery	28/01/2020	28/01/2020
									A071205	Pacemaker Found in Back-Up Mode	28/01/2020	28/01/2020
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	28/01/2020	28/01/2020
									A071205	Pacemaker Found in Back-Up Mode	28/01/2020	28/01/2020
					CT2750	Implantable cardiac electrical stimulators and associated devices	E2311	Discomfort	A070504	Premature Discharge of Battery	28/01/2020	28/01/2020
									A071205	Pacemaker Found in Back-Up Mode	28/01/2020	28/01/2020
							F1901	Additional Surgery	A070504	Premature Discharge of Battery	28/01/2020	28/01/2020
									A071205	Pacemaker Found in Back-Up Mode	28/01/2020	28/01/2020
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	28/01/2020	28/01/2020
									A071205	Pacemaker Found in Back-Up Mode	28/01/2020	28/01/2020
					CT2866	Implantable cardiac electrical stimulators	E2311	Discomfort	A070504	Premature Discharge of Battery	28/01/2020	28/01/2020
									A071205	Pacemaker Found in Back-Up Mode	28/01/2020	28/01/2020
							F1901	Additional Surgery	A070504	Premature Discharge of Battery	28/01/2020	28/01/2020
									A071205	Pacemaker Found in Back-Up Mode	28/01/2020	28/01/2020
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	28/01/2020	28/01/2020
									A071205	Pacemaker Found in Back-Up Mode	28/01/2020	28/01/2020
					CT2985	Active	E2311	Discomfort	A070504	Premature Discharge of Battery	28/01/2020	28/01/2020
									A071205	Pacemaker Found in Back-Up Mode	28/01/2020	28/01/2020
							F1901	Additional Surgery	A070504	Premature Discharge of Battery	28/01/2020	28/01/2020
									A071205	Pacemaker Found in Back-Up Mode	28/01/2020	28/01/2020
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	28/01/2020	28/01/2020
									A071205	Pacemaker Found in Back-Up Mode	28/01/2020	28/01/2020
					CT3001	Implantable metal	E2311	Discomfort	A070504	Premature Discharge of Battery	28/01/2020	28/01/2020
									A071205	Pacemaker Found in Back-Up Mode	28/01/2020	28/01/2020
							F1901	Additional Surgery	A070504	Premature Discharge of Battery	28/01/2020	28/01/2020
									A071205	Pacemaker Found in Back-Up Mode	28/01/2020	28/01/2020
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	28/01/2020	28/01/2020
									A071205	Pacemaker Found in Back-Up Mode	28/01/2020	28/01/2020
					CT3002	Implantable synthetic polymer	E2311	Discomfort	A070504	Premature Discharge of Battery	28/01/2020	28/01/2020
									A071205	Pacemaker Found in Back-Up Mode	28/01/2020	28/01/2020
							F1901	Additional Surgery	A070504	Premature Discharge of Battery	28/01/2020	28/01/2020
									A071205	Pacemaker Found in Back-Up Mode	28/01/2020	28/01/2020

							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	28/01/2020	28/01/2020
									A071205	Pacemaker Found in Back-Up Mode	28/01/2020	28/01/2020
					CT3006	Implantable inorganic material	E2311	Discomfort	A070504	Premature Discharge of Battery	28/01/2020	28/01/2020
									A071205	Pacemaker Found in Back-Up Mode	28/01/2020	28/01/2020
							F1901	Additional Surgery	A070504	Premature Discharge of Battery	28/01/2020	28/01/2020
									A071205	Pacemaker Found in Back-Up Mode	28/01/2020	28/01/2020
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	28/01/2020	28/01/2020
									A071205	Pacemaker Found in Back-Up Mode	28/01/2020	28/01/2020
					CT314	Electrically powered	E2311	Discomfort	A070504	Premature Discharge of Battery	28/01/2020	28/01/2020
									A071205	Pacemaker Found in Back-Up Mode	28/01/2020	28/01/2020
							F1901	Additional Surgery	A070504	Premature Discharge of Battery	28/01/2020	28/01/2020
									A071205	Pacemaker Found in Back-Up Mode	28/01/2020	28/01/2020
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	28/01/2020	28/01/2020
									A071205	Pacemaker Found in Back-Up Mode	28/01/2020	28/01/2020
					CT321	Long-term surgical invasive	E2311	Discomfort	A070504	Premature Discharge of Battery	28/01/2020	28/01/2020
									A071205	Pacemaker Found in Back-Up Mode	28/01/2020	28/01/2020
							F1901	Additional Surgery	A070504	Premature Discharge of Battery	28/01/2020	28/01/2020
									A071205	Pacemaker Found in Back-Up Mode	28/01/2020	28/01/2020
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	28/01/2020	28/01/2020
									A071205	Pacemaker Found in Back-Up Mode	28/01/2020	28/01/2020
					CT449	Generators and associated devices	E2311	Discomfort	A070504	Premature Discharge of Battery	28/01/2020	28/01/2020
									A071205	Pacemaker Found in Back-Up Mode	28/01/2020	28/01/2020
							F1901	Additional Surgery	A070504	Premature Discharge of Battery	28/01/2020	28/01/2020
									A071205	Pacemaker Found in Back-Up Mode	28/01/2020	28/01/2020
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	28/01/2020	28/01/2020
									A071205	Pacemaker Found in Back-Up Mode	28/01/2020	28/01/2020
					CT981	Single use	E2311	Discomfort	A070504	Premature Discharge of Battery	28/01/2020	28/01/2020
									A071205	Pacemaker Found in Back-Up Mode	28/01/2020	28/01/2020
							F1901	Additional Surgery	A070504	Premature Discharge of Battery	28/01/2020	28/01/2020
									A071205	Pacemaker Found in Back-Up Mode	28/01/2020	28/01/2020
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	28/01/2020	28/01/2020
									A071205	Pacemaker Found in Back-Up Mode	28/01/2020	28/01/2020
					CT983	Surgical invasive	E2311	Discomfort	A070504	Premature Discharge of Battery	28/01/2020	28/01/2020
									A071205	Pacemaker Found in Back-Up Mode	28/01/2020	28/01/2020
							F1901	Additional Surgery	A070504	Premature Discharge of Battery	28/01/2020	28/01/2020
									A071205	Pacemaker Found in Back-Up Mode	28/01/2020	28/01/2020
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	28/01/2020	28/01/2020
									A071205	Pacemaker Found in Back-Up Mode	28/01/2020	28/01/2020
8704	2019	All other reportable incidents	35223	Endocardial/interventricular septal pacing lead	CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070101	Failure to Capture		
8705	2020	All other reportable incidents	35039	Epicardial pacing lead	CT1684	Cardiac pacemakers and associated devices	E060104	Bradycardia	A0401	Break		
									A040101	Fracture		
									A070101	Failure to Capture		
									A070102	High Capture Threshold		
									A0712	Pacing Problem		
									A072201	High Impedance		
							E233001	Chest Pain	A0401	Break		
									A040101	Fracture		
									A070101	Failure to Capture		
									A070102	High Capture Threshold		
									A0712	Pacing Problem		
									A072201	High Impedance		
							F1905	Device Revision or Replacement	A0401	Break		
									A040101	Fracture		
									A070101	Failure to Capture		
									A070102	High Capture Threshold		
									A0712	Pacing Problem		
									A072201	High Impedance		
8706	2019	All other reportable incidents	36241	Pacemaker/defibrillator lead	CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A0405	Degraded		
8707	2020	All other reportable incidents	47265	Dual-chamber implantable pacemaker, rate-responsive	CT141	Electrodes/Leads and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT145	Cardiovascular devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT233	Surgical	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT2406	Implantable	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT2407	Active implantable	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT2727	Active implant devices/ancillaries	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT2750	Implantable cardiac electrical stimulators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT2985	Active	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT3001	Implantable metal	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT3002	Implantable synthetic polymer	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT3006	Implantable inorganic material	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		

							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT314	Electrically powered	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT321	Long-term surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT449	Generators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT981	Single use	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT983	Surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
8708	2020	All other reportable incidents	47265	Dual-chamber implantable pacemaker, rate-responsive	CT141	Electrodes/Leads and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	18/02/2020	18/02/2020
							F12	Serious Injury/ Illness/ Impairment	A070504	Premature Discharge of Battery	18/02/2020	18/02/2020
							F1903	Device Explantation	A070504	Premature Discharge of Battery	18/02/2020	18/02/2020
					CT145	Cardiovascular devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	18/02/2020	18/02/2020
							F12	Serious Injury/ Illness/ Impairment	A070504	Premature Discharge of Battery	18/02/2020	18/02/2020
							F1903	Device Explantation	A070504	Premature Discharge of Battery	18/02/2020	18/02/2020
					CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	18/02/2020	18/02/2020
							F12	Serious Injury/ Illness/ Impairment	A070504	Premature Discharge of Battery	18/02/2020	18/02/2020
							F1903	Device Explantation	A070504	Premature Discharge of Battery	18/02/2020	18/02/2020
					CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	18/02/2020	18/02/2020
							F12	Serious Injury/ Illness/ Impairment	A070504	Premature Discharge of Battery	18/02/2020	18/02/2020
							F1903	Device Explantation	A070504	Premature Discharge of Battery	18/02/2020	18/02/2020
					CT233	Surgical	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	18/02/2020	18/02/2020
							F12	Serious Injury/ Illness/ Impairment	A070504	Premature Discharge of Battery	18/02/2020	18/02/2020
							F1903	Device Explantation	A070504	Premature Discharge of Battery	18/02/2020	18/02/2020
					CT2406	Implantable	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	18/02/2020	18/02/2020
							F12	Serious Injury/ Illness/ Impairment	A070504	Premature Discharge of Battery	18/02/2020	18/02/2020
							F1903	Device Explantation	A070504	Premature Discharge of Battery	18/02/2020	18/02/2020
					CT2407	Active implantable	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	18/02/2020	18/02/2020
							F12	Serious Injury/ Illness/ Impairment	A070504	Premature Discharge of Battery	18/02/2020	18/02/2020
							F1903	Device Explantation	A070504	Premature Discharge of Battery	18/02/2020	18/02/2020
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	18/02/2020	18/02/2020
							F12	Serious Injury/ Illness/ Impairment	A070504	Premature Discharge of Battery	18/02/2020	18/02/2020
							F1903	Device Explantation	A070504	Premature Discharge of Battery	18/02/2020	18/02/2020
					CT2727	Active implant devices/ancillaries	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	18/02/2020	18/02/2020
							F12	Serious Injury/ Illness/ Impairment	A070504	Premature Discharge of Battery	18/02/2020	18/02/2020
							F1903	Device Explantation	A070504	Premature Discharge of Battery	18/02/2020	18/02/2020
					CT2750	Implantable cardiac electrical stimulators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	18/02/2020	18/02/2020
							F12	Serious Injury/ Illness/ Impairment	A070504	Premature Discharge of Battery	18/02/2020	18/02/2020
							F1903	Device Explantation	A070504	Premature Discharge of Battery	18/02/2020	18/02/2020
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	18/02/2020	18/02/2020
							F12	Serious Injury/ Illness/ Impairment	A070504	Premature Discharge of Battery	18/02/2020	18/02/2020
							F1903	Device Explantation	A070504	Premature Discharge of Battery	18/02/2020	18/02/2020
					CT2985	Active	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	18/02/2020	18/02/2020
							F12	Serious Injury/ Illness/ Impairment	A070504	Premature Discharge of Battery	18/02/2020	18/02/2020
							F1903	Device Explantation	A070504	Premature Discharge of Battery	18/02/2020	18/02/2020
					CT3001	Implantable metal	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	18/02/2020	18/02/2020
							F12	Serious Injury/ Illness/ Impairment	A070504	Premature Discharge of Battery	18/02/2020	18/02/2020
							F1903	Device Explantation	A070504	Premature Discharge of Battery	18/02/2020	18/02/2020
					CT3002	Implantable synthetic polymer	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	18/02/2020	18/02/2020
							F12	Serious Injury/ Illness/ Impairment	A070504	Premature Discharge of Battery	18/02/2020	18/02/2020
							F1903	Device Explantation	A070504	Premature Discharge of Battery	18/02/2020	18/02/2020
					CT3006	Implantable inorganic material	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	18/02/2020	18/02/2020
							F12	Serious Injury/ Illness/ Impairment	A070504	Premature Discharge of Battery	18/02/2020	18/02/2020
							F1903	Device Explantation	A070504	Premature Discharge of Battery	18/02/2020	18/02/2020
					CT314	Electrically powered	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	18/02/2020	18/02/2020
							F12	Serious Injury/ Illness/ Impairment	A070504	Premature Discharge of Battery	18/02/2020	18/02/2020
							F1903	Device Explantation	A070504	Premature Discharge of Battery	18/02/2020	18/02/2020
					CT321	Long-term surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	18/02/2020	18/02/2020
							F12	Serious Injury/ Illness/ Impairment	A070504	Premature Discharge of Battery	18/02/2020	18/02/2020
							F1903	Device Explantation	A070504	Premature Discharge of Battery	18/02/2020	18/02/2020
					CT449	Generators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	18/02/2020	18/02/2020
							F12	Serious Injury/ Illness/ Impairment	A070504	Premature Discharge of Battery	18/02/2020	18/02/2020
							F1903	Device Explantation	A070504	Premature Discharge of Battery	18/02/2020	18/02/2020
					CT981	Single use	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	18/02/2020	18/02/2020
							F12	Serious Injury/ Illness/ Impairment	A070504	Premature Discharge of Battery	18/02/2020	18/02/2020
							F1903	Device Explantation	A070504	Premature Discharge of Battery	18/02/2020	18/02/2020
					CT983	Surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	18/02/2020	18/02/2020
							F12	Serious Injury/ Illness/ Impairment	A070504	Premature Discharge of Battery	18/02/2020	18/02/2020
							F1903	Device Explantation	A070504	Premature Discharge of Battery	18/02/2020	18/02/2020
8709	2020	All other reportable incidents	35852	Single-chamber implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A070909	Oversensing	24/01/2020	24/01/2020
									A0712	Pacing Problem	24/01/2020	24/01/2020
							F1905	Device Revision or Replacement	A070909	Oversensing	24/01/2020	24/01/2020
									A0712	Pacing Problem	24/01/2020	24/01/2020
8710	2020	All other reportable incidents	47270	Cardiac resynchronization therapy implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A0712	Pacing Problem		
							F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A0712	Pacing Problem		
							F26	No Health Consequences or Impact	A0712	Pacing Problem		
8711	2020	All other reportable incidents	36241	Pacemaker/defibrillator lead	CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070101	Failure to Capture		
									A070909	Oversensing		
									A0712	Pacing Problem		
8712	2019	All other reportable incidents	36241	Pacemaker/defibrillator lead	CT1684	Cardiac pacemakers and associated devices	F14	Prolonged Episode of Care	A072201	High Impedance		
8713	2019	All other reportable incidents	17882	Physiologic-monitoring defibrillation system	CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A0203	Defective Device		
8714	2019	All other reportable incidents	17882	Physiologic-monitoring defibrillation system	CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070803	Failure to Power Up		
							F27	Problem Identified During Non-Clinical Procedure	A070803	Failure to Power Up		
8715	2020	All other reportable incidents	37265	Dual-chamber implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	12/02/2020	12/02/2020
8716	2019	All other reportable incidents	17882	Physiologic-monitoring defibrillation system	CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070803	Failure to Power Up		
8717	2019	All other reportable incidents	37265	Dual-chamber implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	21/03/2019	

8718	2020	All other reportable incidents	35852	Single-chamber implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A13	Communication or Transmission Problem		
8719	2020	All other reportable incidents	47265	Dual-chamber implantable pacemaker, rate-responsive	CT141	Electrodes/Leads and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	31/01/2020	31/01/2020
							F19	Surgical Intervention	A070504	Premature Discharge of Battery	31/01/2020	31/01/2020
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	31/01/2020	31/01/2020
					CT145	Cardiovascular devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	31/01/2020	31/01/2020
							F19	Surgical Intervention	A070504	Premature Discharge of Battery	31/01/2020	31/01/2020
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	31/01/2020	31/01/2020
					CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	31/01/2020	31/01/2020
							F19	Surgical Intervention	A070504	Premature Discharge of Battery	31/01/2020	31/01/2020
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	31/01/2020	31/01/2020
					CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	31/01/2020	31/01/2020
							F19	Surgical Intervention	A070504	Premature Discharge of Battery	31/01/2020	31/01/2020
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	31/01/2020	31/01/2020
					CT233	Surgical	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	31/01/2020	31/01/2020
							F19	Surgical Intervention	A070504	Premature Discharge of Battery	31/01/2020	31/01/2020
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	31/01/2020	31/01/2020
					CT2406	Implantable	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	31/01/2020	31/01/2020
							F19	Surgical Intervention	A070504	Premature Discharge of Battery	31/01/2020	31/01/2020
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	31/01/2020	31/01/2020
					CT2407	Active implantable	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	31/01/2020	31/01/2020
							F19	Surgical Intervention	A070504	Premature Discharge of Battery	31/01/2020	31/01/2020
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	31/01/2020	31/01/2020
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	31/01/2020	31/01/2020
							F19	Surgical Intervention	A070504	Premature Discharge of Battery	31/01/2020	31/01/2020
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	31/01/2020	31/01/2020
					CT2727	Active implant devices/ancillaries	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	31/01/2020	31/01/2020
							F19	Surgical Intervention	A070504	Premature Discharge of Battery	31/01/2020	31/01/2020
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	31/01/2020	31/01/2020
					CT2750	Implantable cardiac electrical stimulators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	31/01/2020	31/01/2020
							F19	Surgical Intervention	A070504	Premature Discharge of Battery	31/01/2020	31/01/2020
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	31/01/2020	31/01/2020
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	31/01/2020	31/01/2020
							F19	Surgical Intervention	A070504	Premature Discharge of Battery	31/01/2020	31/01/2020
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	31/01/2020	31/01/2020
					CT2985	Active	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	31/01/2020	31/01/2020
							F19	Surgical Intervention	A070504	Premature Discharge of Battery	31/01/2020	31/01/2020
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	31/01/2020	31/01/2020
					CT3001	Implantable metal	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	31/01/2020	31/01/2020
							F19	Surgical Intervention	A070504	Premature Discharge of Battery	31/01/2020	31/01/2020
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	31/01/2020	31/01/2020
					CT3002	Implantable synthetic polymer	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	31/01/2020	31/01/2020
							F19	Surgical Intervention	A070504	Premature Discharge of Battery	31/01/2020	31/01/2020
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	31/01/2020	31/01/2020
					CT3006	Implantable inorganic material	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	31/01/2020	31/01/2020
							F19	Surgical Intervention	A070504	Premature Discharge of Battery	31/01/2020	31/01/2020
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	31/01/2020	31/01/2020
					CT314	Electrically powered	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	31/01/2020	31/01/2020
							F19	Surgical Intervention	A070504	Premature Discharge of Battery	31/01/2020	31/01/2020
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	31/01/2020	31/01/2020
					CT321	Long-term surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	31/01/2020	31/01/2020
							F19	Surgical Intervention	A070504	Premature Discharge of Battery	31/01/2020	31/01/2020
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	31/01/2020	31/01/2020
					CT449	Generators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	31/01/2020	31/01/2020
							F19	Surgical Intervention	A070504	Premature Discharge of Battery	31/01/2020	31/01/2020
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	31/01/2020	31/01/2020
					CT981	Single use	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	31/01/2020	31/01/2020
							F19	Surgical Intervention	A070504	Premature Discharge of Battery	31/01/2020	31/01/2020
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	31/01/2020	31/01/2020
					CT983	Surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	31/01/2020	31/01/2020
							F19	Surgical Intervention	A070504	Premature Discharge of Battery	31/01/2020	31/01/2020
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	31/01/2020	31/01/2020
8720	2020	All other reportable incidents	37265	Dual-chamber implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A0722	Impedance Problem		
									A072201	High Impedance		
									A090801	Signal Artifact/Noise		
8721	2020	All other reportable incidents	47265	Dual-chamber implantable pacemaker, rate-responsive	CT141	Electrodes/Leads and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070101	Failure to Capture		
									A0709	Device Sensing Problem		
									A070909	Oversensing		
									A090801	Signal Artifact/Noise		
					CT145	Cardiovascular devices	E2403	No Clinical Signs, Symptoms or Conditions	A070101	Failure to Capture		
									A0709	Device Sensing Problem		
									A070909	Oversensing		
									A090801	Signal Artifact/Noise		
					CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070101	Failure to Capture		
									A0709	Device Sensing Problem		
									A070909	Oversensing		
									A090801	Signal Artifact/Noise		
					CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A070101	Failure to Capture		
									A0709	Device Sensing Problem		
									A070909	Oversensing		
									A090801	Signal Artifact/Noise		
					CT233	Surgical	E2403	No Clinical Signs, Symptoms or Conditions	A070101	Failure to Capture		
									A0709	Device Sensing Problem		
									A070909	Oversensing		
									A090801	Signal Artifact/Noise		
					CT2406	Implantable	E2403	No Clinical Signs, Symptoms or Conditions	A070101	Failure to Capture		
									A0709	Device Sensing Problem		
									A070909	Oversensing		
									A090801	Signal Artifact/Noise		

					CT2407	Active implantable	E2403	No Clinical Signs, Symptoms or Conditions	A070101	Failure to Capture		
									A0709	Device Sensing Problem		
									A070909	Oversensing		
									A090801	Signal Artifact/Noise		
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A070101	Failure to Capture		
									A0709	Device Sensing Problem		
									A070909	Oversensing		
									A090801	Signal Artifact/Noise		
					CT2727	Active implant devices/ancillaries	E2403	No Clinical Signs, Symptoms or Conditions	A070101	Failure to Capture		
									A0709	Device Sensing Problem		
									A070909	Oversensing		
									A090801	Signal Artifact/Noise		
					CT2750	Implantable cardiac electrical stimulators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070101	Failure to Capture		
									A0709	Device Sensing Problem		
									A070909	Oversensing		
									A090801	Signal Artifact/Noise		
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A070101	Failure to Capture		
									A0709	Device Sensing Problem		
									A070909	Oversensing		
									A090801	Signal Artifact/Noise		
					CT2985	Active	E2403	No Clinical Signs, Symptoms or Conditions	A070101	Failure to Capture		
									A0709	Device Sensing Problem		
									A070909	Oversensing		
									A090801	Signal Artifact/Noise		
					CT3001	Implantable metal	E2403	No Clinical Signs, Symptoms or Conditions	A070101	Failure to Capture		
									A0709	Device Sensing Problem		
									A070909	Oversensing		
									A090801	Signal Artifact/Noise		
					CT3002	Implantable synthetic polymer	E2403	No Clinical Signs, Symptoms or Conditions	A070101	Failure to Capture		
									A0709	Device Sensing Problem		
									A070909	Oversensing		
									A090801	Signal Artifact/Noise		
					CT3006	Implantable inorganic material	E2403	No Clinical Signs, Symptoms or Conditions	A070101	Failure to Capture		
									A0709	Device Sensing Problem		
									A070909	Oversensing		
									A090801	Signal Artifact/Noise		
					CT314	Electrically powered	E2403	No Clinical Signs, Symptoms or Conditions	A070101	Failure to Capture		
									A0709	Device Sensing Problem		
									A070909	Oversensing		
									A090801	Signal Artifact/Noise		
					CT321	Long-term surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A070101	Failure to Capture		
									A0709	Device Sensing Problem		
									A070909	Oversensing		
									A090801	Signal Artifact/Noise		
					CT449	Generators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070101	Failure to Capture		
									A0709	Device Sensing Problem		
									A070909	Oversensing		
									A090801	Signal Artifact/Noise		
					CT981	Single use	E2403	No Clinical Signs, Symptoms or Conditions	A070101	Failure to Capture		
									A0709	Device Sensing Problem		
									A070909	Oversensing		
									A090801	Signal Artifact/Noise		
					CT983	Surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A070101	Failure to Capture		
									A0709	Device Sensing Problem		
									A070909	Oversensing		
									A090801	Signal Artifact/Noise		
8722	2020	All other reportable incidents	36241	Pacemaker/defibrillator lead	CT1684	Cardiac pacemakers and associated devices	F21	Unexpected Deterioration	A040101	Fracture		
									A070101	Failure to Capture		
									A070102	High Capture Threshold		
									A072201	High Impedance		
8723	2020	All other reportable incidents	37265	Dual-chamber implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	27/01/2020	27/01/2020
8724	2020	All other reportable incidents	37265	Dual-chamber implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	07/02/2020	07/02/2020
							F19	Surgical Intervention	A070504	Premature Discharge of Battery	07/02/2020	07/02/2020
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	07/02/2020	07/02/2020
8725	2020	All other reportable incidents	47265	Dual-chamber implantable pacemaker, rate-responsive	CT141	Electrodes/Leads and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070101	Failure to Capture		
							F26	No Health Consequences or Impact	A070101	Failure to Capture		
					CT145	Cardiovascular devices	E2403	No Clinical Signs, Symptoms or Conditions	A070101	Failure to Capture		
							F26	No Health Consequences or Impact	A070101	Failure to Capture		
					CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070101	Failure to Capture		
							F26	No Health Consequences or Impact	A070101	Failure to Capture		
					CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A070101	Failure to Capture		
							F26	No Health Consequences or Impact	A070101	Failure to Capture		
					CT233	Surgical	E2403	No Clinical Signs, Symptoms or Conditions	A070101	Failure to Capture		
							F26	No Health Consequences or Impact	A070101	Failure to Capture		
					CT2406	Implantable	E2403	No Clinical Signs, Symptoms or Conditions	A070101	Failure to Capture		
							F26	No Health Consequences or Impact	A070101	Failure to Capture		
					CT2407	Active implantable	E2403	No Clinical Signs, Symptoms or Conditions	A070101	Failure to Capture		
							F26	No Health Consequences or Impact	A070101	Failure to Capture		
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A070101	Failure to Capture		
							F26	No Health Consequences or Impact	A070101	Failure to Capture		
					CT2727	Active implant devices/ancillaries	E2403	No Clinical Signs, Symptoms or Conditions	A070101	Failure to Capture		
							F26	No Health Consequences or Impact	A070101	Failure to Capture		
					CT2750	Implantable cardiac electrical stimulators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070101	Failure to Capture		
							F26	No Health Consequences or Impact	A070101	Failure to Capture		
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A070101	Failure to Capture		
							F26	No Health Consequences or Impact	A070101	Failure to Capture		
					CT2985	Active	E2403	No Clinical Signs, Symptoms or Conditions	A070101	Failure to Capture		
							F26	No Health Consequences or Impact	A070101	Failure to Capture		

					CT3001	Implantable metal	E2403	No Clinical Signs, Symptoms or Conditions	A070101	Failure to Capture		
							F26	No Health Consequences or Impact	A070101	Failure to Capture		
					CT3002	Implantable synthetic polymer	E2403	No Clinical Signs, Symptoms or Conditions	A070101	Failure to Capture		
							F26	No Health Consequences or Impact	A070101	Failure to Capture		
					CT3006	Implantable inorganic material	E2403	No Clinical Signs, Symptoms or Conditions	A070101	Failure to Capture		
							F26	No Health Consequences or Impact	A070101	Failure to Capture		
					CT314	Electrically powered	E2403	No Clinical Signs, Symptoms or Conditions	A070101	Failure to Capture		
							F26	No Health Consequences or Impact	A070101	Failure to Capture		
					CT321	Long-term surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A070101	Failure to Capture		
							F26	No Health Consequences or Impact	A070101	Failure to Capture		
					CT449	Generators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070101	Failure to Capture		
							F26	No Health Consequences or Impact	A070101	Failure to Capture		
					CT981	Single use	E2403	No Clinical Signs, Symptoms or Conditions	A070101	Failure to Capture		
							F26	No Health Consequences or Impact	A070101	Failure to Capture		
					CT983	Surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A070101	Failure to Capture		
							F26	No Health Consequences or Impact	A070101	Failure to Capture		
8726	2020	All other reportable incidents	36241	Pacemaker/defibrillator lead	CT1684	Cardiac pacemakers and associated devices	F14	Prolonged Episode of Care	A051201	Device Dislodged or Dislocated	15/01/2020	15/01/2020
8727	2020	All other reportable incidents	36241	Pacemaker/defibrillator lead	CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A072201	High Impedance		
8728	2020	All other reportable incidents	35852	Single-chamber implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E2104	Electric Shock	A070909	Oversensing		
									A071302	Inappropriate/Inadequate Shock/Stimulation		
							E2403	No Clinical Signs, Symptoms or Conditions	A070909	Oversensing		
									A071302	Inappropriate/Inadequate Shock/Stimulation		
							F26	No Health Consequences or Impact	A070909	Oversensing		
									A071302	Inappropriate/Inadequate Shock/Stimulation		
8729	2019	All other reportable incidents	47265	Dual-chamber implantable pacemaker, rate-responsive	CT141	Electrodes/Leads and associated devices	F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A110206	Application Program Version or Upgrade Problem		
					CT145	Cardiovascular devices	F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A110206	Application Program Version or Upgrade Problem		
					CT1684	Cardiac pacemakers and associated devices	F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A110206	Application Program Version or Upgrade Problem		
					CT1685	Cardiac pacemakers	F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A110206	Application Program Version or Upgrade Problem		
					CT233	Surgical	F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A110206	Application Program Version or Upgrade Problem		
					CT2406	Implantable	F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A110206	Application Program Version or Upgrade Problem		
					CT2407	Active implantable	F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A110206	Application Program Version or Upgrade Problem		
					CT257	Implantable cardiac pacemakers	F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A110206	Application Program Version or Upgrade Problem		
					CT2727	Active implant devices/ancillaries	F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A110206	Application Program Version or Upgrade Problem		
					CT2750	Implantable cardiac electrical stimulators and associated devices	F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A110206	Application Program Version or Upgrade Problem		
					CT2866	Implantable cardiac electrical stimulators	F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A110206	Application Program Version or Upgrade Problem		
					CT2985	Active	F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A110206	Application Program Version or Upgrade Problem		
					CT3001	Implantable metal	F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A110206	Application Program Version or Upgrade Problem		
					CT3002	Implantable synthetic polymer	F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A110206	Application Program Version or Upgrade Problem		
					CT3006	Implantable inorganic material	F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A110206	Application Program Version or Upgrade Problem		
					CT314	Electrically powered	F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A110206	Application Program Version or Upgrade Problem		
					CT321	Long-term surgical invasive	F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A110206	Application Program Version or Upgrade Problem		
					CT449	Generators and associated devices	F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A110206	Application Program Version or Upgrade Problem		
					CT981	Single use	F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A110206	Application Program Version or Upgrade Problem		
					CT983	Surgical invasive	F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A110206	Application Program Version or Upgrade Problem		
8730	2019	All other reportable incidents	47265	Dual-chamber implantable pacemaker, rate-responsive	CT141	Electrodes/Leads and associated devices	F1905	Device Revision or Replacement	A160601	Premature Elective Replacement Indicator	29/03/2019	
					CT145	Cardiovascular devices	F1905	Device Revision or Replacement	A160601	Premature Elective Replacement Indicator	29/03/2019	
					CT1684	Cardiac pacemakers and associated devices	F1905	Device Revision or Replacement	A160601	Premature Elective Replacement Indicator	29/03/2019	
					CT1685	Cardiac pacemakers	F1905	Device Revision or Replacement	A160601	Premature Elective Replacement Indicator	29/03/2019	
					CT233	Surgical	F1905	Device Revision or Replacement	A160601	Premature Elective Replacement Indicator	29/03/2019	
					CT2406	Implantable	F1905	Device Revision or Replacement	A160601	Premature Elective Replacement Indicator	29/03/2019	
					CT2407	Active implantable	F1905	Device Revision or Replacement	A160601	Premature Elective Replacement Indicator	29/03/2019	
					CT257	Implantable cardiac pacemakers	F1905	Device Revision or Replacement	A160601	Premature Elective Replacement Indicator	29/03/2019	
					CT2727	Active implant devices/ancillaries	F1905	Device Revision or Replacement	A160601	Premature Elective Replacement Indicator	29/03/2019	
					CT2750	Implantable cardiac electrical stimulators and associated devices	F1905	Device Revision or Replacement	A160601	Premature Elective Replacement Indicator	29/03/2019	
					CT2866	Implantable cardiac electrical stimulators	F1905	Device Revision or Replacement	A160601	Premature Elective Replacement Indicator	29/03/2019	
					CT2985	Active	F1905	Device Revision or Replacement	A160601	Premature Elective Replacement Indicator	29/03/2019	
					CT3001	Implantable metal	F1905	Device Revision or Replacement	A160601	Premature Elective Replacement Indicator	29/03/2019	
					CT3002	Implantable synthetic polymer	F1905	Device Revision or Replacement	A160601	Premature Elective Replacement Indicator	29/03/2019	
					CT3006	Implantable inorganic material	F1905	Device Revision or Replacement	A160601	Premature Elective Replacement Indicator	29/03/2019	
					CT314	Electrically powered	F1905	Device Revision or Replacement	A160601	Premature Elective Replacement Indicator	29/03/2019	
					CT321	Long-term surgical invasive	F1905	Device Revision or Replacement	A160601	Premature Elective Replacement Indicator	29/03/2019	
					CT449	Generators and associated devices	F1905	Device Revision or Replacement	A160601	Premature Elective Replacement Indicator	29/03/2019	
					CT981	Single use	F1905	Device Revision or Replacement	A160601	Premature Elective Replacement Indicator	29/03/2019	
					CT983	Surgical invasive	F1905	Device Revision or Replacement	A160601	Premature Elective Replacement Indicator	29/03/2019	
8731	2019	All other reportable incidents	35852	Single-chamber implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem	11/04/2019	
8732	2019	All other reportable incidents	35852	Single-chamber implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A071302	Inappropriate/Inadequate Shock/Stimulation		
8733	2019	All other reportable incidents	36241	Pacemaker/defibrillator lead	CT1684	Cardiac pacemakers and associated devices	F1905	Device Revision or Replacement	A0510	Retraction Problem	18/03/2019	
8734	2020	All other reportable incidents	36241	Pacemaker/defibrillator lead	CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070909	Oversensing		
8735	2019	All other reportable incidents	35852	Single-chamber implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A070909	Oversensing		
8736	2020	All other reportable incidents	47270	Cardiac resynchronization therapy implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A072201	High Impedance		
8737	2020	All other reportable incidents	36241	Pacemaker/defibrillator lead	CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A040101	Fracture	16/01/2020	16/01/2020
									A05	Mechanical Problem	16/01/2020	16/01/2020
									A072201	High Impedance	16/01/2020	16/01/2020
							F08	Hospitalization or Prolonged Hospitalization	A040101	Fracture	16/01/2020	16/01/2020
									A05	Mechanical Problem	16/01/2020	16/01/2020
									A072201	High Impedance	16/01/2020	16/01/2020
							F19	Surgical Intervention	A040101	Fracture	16/01/2020	16/01/2020
									A05	Mechanical Problem	16/01/2020	16/01/2020
									A072201	High Impedance	16/01/2020	16/01/2020
8738	2020	All other reportable incidents	35852	Single-chamber implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A072201	High Impedance		
							F26	No Health Consequences or Impact	A072201	High Impedance		
8739	2020	All other reportable incidents	37265	Dual-chamber implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	12/02/2020	12/02/2020
8740	2020	All other reportable incidents	37265	Dual-chamber implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E0717	Dyspnea	A071302	Inappropriate/Inadequate Shock/Stimulation		
							E2403	No Clinical Signs, Symptoms or Conditions	A071302	Inappropriate/Inadequate Shock/Stimulation		
							F26	No Health Consequences or Impact	A071302	Inappropriate/Inadequate Shock/Stimulation		

8741	2020	All other reportable incidents	36241	Pacemaker/defibrillator lead	CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070909	Oversensing		
8742	2020	All other reportable incidents	17882	Physiologic-monitoring defibrillation system	CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070908	Failure to Sense		
							F21	Unexpected Deterioration	A070908	Failure to Sense		
							F26	No Health Consequences or Impact	A070908	Failure to Sense		
8743	2019	All other reportable incidents	36241	Pacemaker/defibrillator lead	CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A072201	High Impedance	12/04/2021	12/04/2021
							F19	Surgical Intervention	A072201	High Impedance	12/04/2021	12/04/2021
8744	2019	All other reportable incidents	47265	Dual-chamber implantable pacemaker, rate-responsive	CT141	Electrodes/Leads and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A0710	Failure to Conduct		
					CT145	Cardiovascular devices	E2403	No Clinical Signs, Symptoms or Conditions	A0710	Failure to Conduct		
					CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A0710	Failure to Conduct		
					CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A0710	Failure to Conduct		
					CT233	Surgical	E2403	No Clinical Signs, Symptoms or Conditions	A0710	Failure to Conduct		
					CT2406	Implantable	E2403	No Clinical Signs, Symptoms or Conditions	A0710	Failure to Conduct		
					CT2407	Active implantable	E2403	No Clinical Signs, Symptoms or Conditions	A0710	Failure to Conduct		
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A0710	Failure to Conduct		
					CT2727	Active implant devices/ancillaries	E2403	No Clinical Signs, Symptoms or Conditions	A0710	Failure to Conduct		
					CT2750	Implantable cardiac electrical stimulators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A0710	Failure to Conduct		
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A0710	Failure to Conduct		
					CT2985	Active	E2403	No Clinical Signs, Symptoms or Conditions	A0710	Failure to Conduct		
					CT3001	Implantable metal	E2403	No Clinical Signs, Symptoms or Conditions	A0710	Failure to Conduct		
					CT3002	Implantable synthetic polymer	E2403	No Clinical Signs, Symptoms or Conditions	A0710	Failure to Conduct		
					CT3006	Implantable inorganic material	E2403	No Clinical Signs, Symptoms or Conditions	A0710	Failure to Conduct		
					CT314	Electrically powered	E2403	No Clinical Signs, Symptoms or Conditions	A0710	Failure to Conduct		
					CT321	Long-term surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A0710	Failure to Conduct		
					CT449	Generators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A0710	Failure to Conduct		
					CT981	Single use	E2403	No Clinical Signs, Symptoms or Conditions	A0710	Failure to Conduct		
					CT983	Surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A0710	Failure to Conduct		
8745	2019	All other reportable incidents	37265	Dual-chamber implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A070504	Premature Discharge of Battery	26/04/2019	
8746	2019	All other reportable incidents	36241	Pacemaker/defibrillator lead	CT1684	Cardiac pacemakers and associated devices	F1905	Device Revision or Replacement	A072201	High Impedance		
8747	2019	All other reportable incidents	35852	Single-chamber implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A070909	Oversensing		
8748	2019	All other reportable incidents	37265	Dual-chamber implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	F1905	Device Revision or Replacement	A071205	Pacemaker Found in Back-Up Mode	04/04/2019	
8749	2019	All other reportable incidents	47265	Dual-chamber implantable pacemaker, rate-responsive	CT141	Electrodes/Leads and associated devices	F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A070101	Failure to Capture		
					CT145	Cardiovascular devices	F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A070101	Failure to Capture		
					CT1684	Cardiac pacemakers and associated devices	F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A070101	Failure to Capture		
					CT1685	Cardiac pacemakers	F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A070101	Failure to Capture		
					CT233	Surgical	F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A070101	Failure to Capture		
					CT2406	Implantable	F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A070101	Failure to Capture		
					CT2407	Active implantable	F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A070101	Failure to Capture		
					CT257	Implantable cardiac pacemakers	F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A070101	Failure to Capture		
					CT2727	Active implant devices/ancillaries	F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A070101	Failure to Capture		
					CT2750	Implantable cardiac electrical stimulators and associated devices	F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A070101	Failure to Capture		
					CT2866	Implantable cardiac electrical stimulators	F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A070101	Failure to Capture		
					CT2985	Active	F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A070101	Failure to Capture		
					CT3001	Implantable metal	F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A070101	Failure to Capture		
					CT3002	Implantable synthetic polymer	F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A070101	Failure to Capture		
					CT3006	Implantable inorganic material	F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A070101	Failure to Capture		
					CT314	Electrically powered	F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A070101	Failure to Capture		
					CT321	Long-term surgical invasive	F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A070101	Failure to Capture		
					CT449	Generators and associated devices	F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A070101	Failure to Capture		
					CT981	Single use	F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A070101	Failure to Capture		
					CT983	Surgical invasive	F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A070101	Failure to Capture		
8750	2019	All other reportable incidents	36241	Pacemaker/defibrillator lead	CT1684	Cardiac pacemakers and associated devices	E2109	Implant Pain	A04	Material Integrity Problem	20/03/2019	
8751	2020	All other reportable incidents	17882	Physiologic-monitoring defibrillation system	CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070803	Failure to Power Up		
							F27	Problem Identified During Non-Clinical Procedure	A070803	Failure to Power Up		
8752	2019	All other reportable incidents	35852	Single-chamber implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery		26/04/2019
8753	2019	All other reportable incidents	60789	Intracardiac pacemaker	CT130	Catheters and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A0712	Pacing Problem		
					CT145	Cardiovascular devices	E2403	No Clinical Signs, Symptoms or Conditions	A0712	Pacing Problem		
					CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A0712	Pacing Problem		
					CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A0712	Pacing Problem		
					CT233	Surgical	E2403	No Clinical Signs, Symptoms or Conditions	A0712	Pacing Problem		
					CT2406	Implantable	E2403	No Clinical Signs, Symptoms or Conditions	A0712	Pacing Problem		
					CT2407	Active implantable	E2403	No Clinical Signs, Symptoms or Conditions	A0712	Pacing Problem		
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A0712	Pacing Problem		
					CT2727	Active implant devices/ancillaries	E2403	No Clinical Signs, Symptoms or Conditions	A0712	Pacing Problem		
					CT2750	Implantable cardiac electrical stimulators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A0712	Pacing Problem		
					CT2818	Catheterization/Cannulation/Fluid path devices	E2403	No Clinical Signs, Symptoms or Conditions	A0712	Pacing Problem		
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A0712	Pacing Problem		
					CT2985	Active	E2403	No Clinical Signs, Symptoms or Conditions	A0712	Pacing Problem		
					CT3001	Implantable metal	E2403	No Clinical Signs, Symptoms or Conditions	A0712	Pacing Problem		
					CT3002	Implantable synthetic polymer	E2403	No Clinical Signs, Symptoms or Conditions	A0712	Pacing Problem		
					CT3006	Implantable inorganic material	E2403	No Clinical Signs, Symptoms or Conditions	A0712	Pacing Problem		
					CT314	Electrically powered	E2403	No Clinical Signs, Symptoms or Conditions	A0712	Pacing Problem		
					CT321	Long-term surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A0712	Pacing Problem		
					CT981	Single use	E2403	No Clinical Signs, Symptoms or Conditions	A0712	Pacing Problem		
					CT983	Surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A0712	Pacing Problem		
8754	2020	All other reportable incidents	36241	Pacemaker/defibrillator lead	CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A05	Mechanical Problem		
									A0709	Device Sensing Problem		
									A22	Human-Device Interface Problem		
8755	2020	All other reportable incidents	47267	Single-chamber implantable pacemaker, rate-responsive	CT141	Electrodes/Leads and associated devices	F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery		
					CT145	Cardiovascular devices	F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery		
					CT1684	Cardiac pacemakers and associated devices	F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery		
					CT1685	Cardiac pacemakers	F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery		
					CT233	Surgical	F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery		
					CT2406	Implantable	F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery		
					CT2407	Active implantable	F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery		
					CT257	Implantable cardiac pacemakers	F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery		

					CT2727	Active implant devices/ancillaries	F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery		
					CT2750	Implantable cardiac electrical stimulators and associated devices	F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery		
					CT2866	Implantable cardiac electrical stimulators	F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery		
					CT2985	Active	F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery		
					CT3001	Implantable metal	F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery		
					CT3002	Implantable synthetic polymer	F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery		
					CT3006	Implantable inorganic material	F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery		
					CT314	Electrically powered	F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery		
					CT321	Long-term surgical invasive	F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery		
					CT449	Generators and associated devices	F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery		
					CT981	Single use	F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery		
					CT983	Surgical invasive	F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery		
8756	2020	All other reportable incidents	47265	Dual-chamber implantable pacemaker, rate-responsive	CT141	Electrodes/Leads and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	17/03/2020	17/03/2020
							F12	Serious Injury/ Illness/ Impairment	A070504	Premature Discharge of Battery	17/03/2020	17/03/2020
							F1903	Device Explantation	A070504	Premature Discharge of Battery	17/03/2020	17/03/2020
					CT145	Cardiovascular devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	17/03/2020	17/03/2020
							F12	Serious Injury/ Illness/ Impairment	A070504	Premature Discharge of Battery	17/03/2020	17/03/2020
							F1903	Device Explantation	A070504	Premature Discharge of Battery	17/03/2020	17/03/2020
					CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	17/03/2020	17/03/2020
							F12	Serious Injury/ Illness/ Impairment	A070504	Premature Discharge of Battery	17/03/2020	17/03/2020
							F1903	Device Explantation	A070504	Premature Discharge of Battery	17/03/2020	17/03/2020
					CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	17/03/2020	17/03/2020
							F12	Serious Injury/ Illness/ Impairment	A070504	Premature Discharge of Battery	17/03/2020	17/03/2020
							F1903	Device Explantation	A070504	Premature Discharge of Battery	17/03/2020	17/03/2020
					CT233	Surgical	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	17/03/2020	17/03/2020
							F12	Serious Injury/ Illness/ Impairment	A070504	Premature Discharge of Battery	17/03/2020	17/03/2020
							F1903	Device Explantation	A070504	Premature Discharge of Battery	17/03/2020	17/03/2020
					CT2406	Implantable	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	17/03/2020	17/03/2020
							F12	Serious Injury/ Illness/ Impairment	A070504	Premature Discharge of Battery	17/03/2020	17/03/2020
							F1903	Device Explantation	A070504	Premature Discharge of Battery	17/03/2020	17/03/2020
					CT2407	Active implantable	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	17/03/2020	17/03/2020
							F12	Serious Injury/ Illness/ Impairment	A070504	Premature Discharge of Battery	17/03/2020	17/03/2020
							F1903	Device Explantation	A070504	Premature Discharge of Battery	17/03/2020	17/03/2020
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	17/03/2020	17/03/2020
							F12	Serious Injury/ Illness/ Impairment	A070504	Premature Discharge of Battery	17/03/2020	17/03/2020
							F1903	Device Explantation	A070504	Premature Discharge of Battery	17/03/2020	17/03/2020
					CT2727	Active implant devices/ancillaries	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	17/03/2020	17/03/2020
							F12	Serious Injury/ Illness/ Impairment	A070504	Premature Discharge of Battery	17/03/2020	17/03/2020
							F1903	Device Explantation	A070504	Premature Discharge of Battery	17/03/2020	17/03/2020
					CT2750	Implantable cardiac electrical stimulators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	17/03/2020	17/03/2020
							F12	Serious Injury/ Illness/ Impairment	A070504	Premature Discharge of Battery	17/03/2020	17/03/2020
							F1903	Device Explantation	A070504	Premature Discharge of Battery	17/03/2020	17/03/2020
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	17/03/2020	17/03/2020
							F12	Serious Injury/ Illness/ Impairment	A070504	Premature Discharge of Battery	17/03/2020	17/03/2020
							F1903	Device Explantation	A070504	Premature Discharge of Battery	17/03/2020	17/03/2020
					CT2985	Active	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	17/03/2020	17/03/2020
							F12	Serious Injury/ Illness/ Impairment	A070504	Premature Discharge of Battery	17/03/2020	17/03/2020
							F1903	Device Explantation	A070504	Premature Discharge of Battery	17/03/2020	17/03/2020
					CT3001	Implantable metal	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	17/03/2020	17/03/2020
							F12	Serious Injury/ Illness/ Impairment	A070504	Premature Discharge of Battery	17/03/2020	17/03/2020
							F1903	Device Explantation	A070504	Premature Discharge of Battery	17/03/2020	17/03/2020
					CT3002	Implantable synthetic polymer	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	17/03/2020	17/03/2020
							F12	Serious Injury/ Illness/ Impairment	A070504	Premature Discharge of Battery	17/03/2020	17/03/2020
							F1903	Device Explantation	A070504	Premature Discharge of Battery	17/03/2020	17/03/2020
					CT3006	Implantable inorganic material	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	17/03/2020	17/03/2020
							F12	Serious Injury/ Illness/ Impairment	A070504	Premature Discharge of Battery	17/03/2020	17/03/2020
							F1903	Device Explantation	A070504	Premature Discharge of Battery	17/03/2020	17/03/2020
					CT314	Electrically powered	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	17/03/2020	17/03/2020
							F12	Serious Injury/ Illness/ Impairment	A070504	Premature Discharge of Battery	17/03/2020	17/03/2020
							F1903	Device Explantation	A070504	Premature Discharge of Battery	17/03/2020	17/03/2020
					CT321	Long-term surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	17/03/2020	17/03/2020
							F12	Serious Injury/ Illness/ Impairment	A070504	Premature Discharge of Battery	17/03/2020	17/03/2020
							F1903	Device Explantation	A070504	Premature Discharge of Battery	17/03/2020	17/03/2020
					CT449	Generators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	17/03/2020	17/03/2020
							F12	Serious Injury/ Illness/ Impairment	A070504	Premature Discharge of Battery	17/03/2020	17/03/2020
							F1903	Device Explantation	A070504	Premature Discharge of Battery	17/03/2020	17/03/2020
					CT981	Single use	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	17/03/2020	17/03/2020
							F12	Serious Injury/ Illness/ Impairment	A070504	Premature Discharge of Battery	17/03/2020	17/03/2020
							F1903	Device Explantation	A070504	Premature Discharge of Battery	17/03/2020	17/03/2020
					CT983	Surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	17/03/2020	17/03/2020
							F12	Serious Injury/ Illness/ Impairment	A070504	Premature Discharge of Battery	17/03/2020	17/03/2020
							F1903	Device Explantation	A070504	Premature Discharge of Battery	17/03/2020	17/03/2020
8757	2020	All other reportable incidents	35852	Single-chamber implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A070909	Oversensing		
8758	2019	All other reportable incidents	17882	Physiologic-monitoring defibrillation system	CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A090206	No Display/Image		
8759	2019	All other reportable incidents	17882	Physiologic-monitoring defibrillation system	CT1684	Cardiac pacemakers and associated devices	F05	Delay to Treatment/ Therapy	A070908	Failure to Sense		
8760	2019	All other reportable incidents	17882	Physiologic-monitoring defibrillation system	CT1684	Cardiac pacemakers and associated devices	F14	Prolonged Episode of Care	A071301	Failure to Deliver Shock/Stimulation		
8761	2019	All other reportable incidents	35223	Endocardial/interventricular septal pacing lead	CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070909	Oversensing		
8762	2019	All other reportable incidents	47852	Temporary cardiac pacing catheter	CT1684	Cardiac pacemakers and associated devices	E2114	Perforation	A0512	Unintended Movement		
8763	2019	All other reportable incidents	47270	Cardiac resynchronization therapy implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	F1905	Device Revision or Replacement	A071205	Pacemaker Found in Back-Up Mode	10/04/2019	
8764	2019	Death	47270	Cardiac resynchronization therapy implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E2401	Insufficient Information	A26	Insufficient Device Problem Information		
8765	2020	All other reportable incidents	36241	Pacemaker/defibrillator lead	CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A040101	Fracture	07/02/2020	07/02/2020
									A05	Mechanical Problem	07/02/2020	07/02/2020
									A0709	Device Sensing Problem	07/02/2020	07/02/2020
							F26	No Health Consequences or Impact	A040101	Fracture	07/02/2020	07/02/2020
									A05	Mechanical Problem	07/02/2020	07/02/2020

									A0709	Device Sensing Problem	07/02/2020	07/02/2020
8766	2020	All other reportable incidents	37265	Dual-chamber implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	12/12/2019	12/12/2019
									A090201	Device Displays Incorrect Message	12/12/2019	12/12/2019
8767	2020	All other reportable incidents	37265	Dual-chamber implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	02/01/2020	02/01/2020
									A090201	Device Displays Incorrect Message	02/01/2020	02/01/2020
8768	2020	All other reportable incidents	37265	Dual-chamber implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A072202	Low Impedance		
							F26	No Health Consequences or Impact	A072202	Low Impedance		
8769	2020	All other reportable incidents	37265	Dual-chamber implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A0712	Pacing Problem		
8770	2020	All other reportable incidents	37265	Dual-chamber implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A0712	Pacing Problem		
8771	2020	All other reportable incidents	47263	Cardiac resynchronization therapy implantable pacemaker	CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070909	Oversensing		
									A071209	Extracardiac Stimulation		
									A072201	High Impedance		
					CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A070909	Oversensing		
									A071209	Extracardiac Stimulation		
									A072201	High Impedance		
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A070909	Oversensing		
									A071209	Extracardiac Stimulation		
									A072201	High Impedance		
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A070909	Oversensing		
									A071209	Extracardiac Stimulation		
									A072201	High Impedance		
8772	2020	All other reportable incidents	35852	Single-chamber implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	04/02/2020	04/02/2020
									A070602	Delayed Charge Time	04/02/2020	04/02/2020
									A0908	Incorrect, Inadequate or Imprecise Result or Readings	04/02/2020	04/02/2020
8773	2019	All other reportable incidents	47270	Cardiac resynchronization therapy implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A0709	Device Sensing Problem		
8774	2019	All other reportable incidents	37265	Dual-chamber implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	F1905	Device Revision or Replacement	A071302	Inappropriate/Inadequate Shock/Stimulation		
8775	2019	All other reportable incidents	37265	Dual-chamber implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
8776	2019	All other reportable incidents	47270	Cardiac resynchronization therapy implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	01/04/2019	
8777	2019	All other reportable incidents	47265	Dual-chamber implantable pacemaker, rate-responsive	CT141	Electrodes/Leads and associated devices	F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery		
					CT145	Cardiovascular devices	F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery		
					CT1684	Cardiac pacemakers and associated devices	F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery		
					CT1685	Cardiac pacemakers	F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery		
					CT233	Surgical	F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery		
					CT2406	Implantable	F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery		
					CT2407	Active implantable	F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery		
					CT257	Implantable cardiac pacemakers	F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery		
					CT2727	Active implant devices/ancillaries	F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery		
					CT2750	Implantable cardiac electrical stimulators and associated devices	F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery		
					CT2866	Implantable cardiac electrical stimulators	F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery		
					CT2985	Active	F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery		
					CT3001	Implantable metal	F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery		
					CT3002	Implantable synthetic polymer	F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery		
					CT3006	Implantable inorganic material	F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery		
					CT314	Electrically powered	F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery		
					CT321	Long-term surgical invasive	F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery		
					CT449	Generators and associated devices	F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery		
					CT981	Single use	F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery		
					CT983	Surgical invasive	F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery		
8778	2019	All other reportable incidents	35852	Single-chamber implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A072201	High Impedance		
8779	2020	All other reportable incidents	47263	Cardiac resynchronization therapy implantable pacemaker	CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	24/02/2020	24/02/2020
							F1901	Additional Surgery	A070504	Premature Discharge of Battery	24/02/2020	24/02/2020
					CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	24/02/2020	24/02/2020
							F1901	Additional Surgery	A070504	Premature Discharge of Battery	24/02/2020	24/02/2020
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	24/02/2020	24/02/2020
							F1901	Additional Surgery	A070504	Premature Discharge of Battery	24/02/2020	24/02/2020
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	24/02/2020	24/02/2020
							F1901	Additional Surgery	A070504	Premature Discharge of Battery	24/02/2020	24/02/2020
8780	2020	All other reportable incidents	47265	Dual-chamber implantable pacemaker, rate-responsive	CT141	Electrodes/Leads and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070909	Oversensing		
									A072201	High Impedance		
									A090801	Signal Artifact/Noise		
					CT145	Cardiovascular devices	E2403	No Clinical Signs, Symptoms or Conditions	A070909	Oversensing		
									A072201	High Impedance		
									A090801	Signal Artifact/Noise		
					CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070909	Oversensing		
									A072201	High Impedance		
									A090801	Signal Artifact/Noise		
					CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A070909	Oversensing		
									A072201	High Impedance		
									A090801	Signal Artifact/Noise		
					CT233	Surgical	E2403	No Clinical Signs, Symptoms or Conditions	A070909	Oversensing		
									A072201	High Impedance		
									A090801	Signal Artifact/Noise		
					CT2406	Implantable	E2403	No Clinical Signs, Symptoms or Conditions	A070909	Oversensing		
									A072201	High Impedance		
									A090801	Signal Artifact/Noise		
					CT2407	Active implantable	E2403	No Clinical Signs, Symptoms or Conditions	A070909	Oversensing		
									A072201	High Impedance		
									A090801	Signal Artifact/Noise		
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A070909	Oversensing		
									A072201	High Impedance		
									A090801	Signal Artifact/Noise		
					CT2727	Active implant devices/ancillaries	E2403	No Clinical Signs, Symptoms or Conditions	A070909	Oversensing		
									A072201	High Impedance		
									A090801	Signal Artifact/Noise		
					CT2750	Implantable cardiac electrical stimulators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070909	Oversensing		

									A072201	High Impedance		
									A090801	Signal Artifact/Noise		
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A070909	Oversensing		
									A072201	High Impedance		
									A090801	Signal Artifact/Noise		
					CT2985	Active	E2403	No Clinical Signs, Symptoms or Conditions	A070909	Oversensing		
									A072201	High Impedance		
									A090801	Signal Artifact/Noise		
					CT3001	Implantable metal	E2403	No Clinical Signs, Symptoms or Conditions	A070909	Oversensing		
									A072201	High Impedance		
									A090801	Signal Artifact/Noise		
					CT3002	Implantable synthetic polymer	E2403	No Clinical Signs, Symptoms or Conditions	A070909	Oversensing		
									A072201	High Impedance		
									A090801	Signal Artifact/Noise		
					CT3006	Implantable inorganic material	E2403	No Clinical Signs, Symptoms or Conditions	A070909	Oversensing		
									A072201	High Impedance		
									A090801	Signal Artifact/Noise		
					CT314	Electrically powered	E2403	No Clinical Signs, Symptoms or Conditions	A070909	Oversensing		
									A072201	High Impedance		
									A090801	Signal Artifact/Noise		
					CT321	Long-term surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A070909	Oversensing		
									A072201	High Impedance		
									A090801	Signal Artifact/Noise		
					CT449	Generators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070909	Oversensing		
									A072201	High Impedance		
									A090801	Signal Artifact/Noise		
					CT981	Single use	E2403	No Clinical Signs, Symptoms or Conditions	A070909	Oversensing		
									A072201	High Impedance		
									A090801	Signal Artifact/Noise		
					CT983	Surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A070909	Oversensing		
									A072201	High Impedance		
									A090801	Signal Artifact/Noise		
8781	2020	All other reportable incidents	17882	Physiologic-monitoring defibrillation system	CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070802	Intermittent Loss of Power		
							F05	Delay to Treatment/ Therapy	A070802	Intermittent Loss of Power		
							F26	No Health Consequences or Impact	A070802	Intermittent Loss of Power		
8782	2020	All other reportable incidents	17882	Physiologic-monitoring defibrillation system	CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070908	Failure to Sense		
							F27	Problem Identified During Non-Clinical Procedure	A070908	Failure to Sense		
8783	2019	All other reportable incidents	35852	Single-chamber implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A071302	Inappropriate/Inadequate Shock/Stimulation		
8784	2019	All other reportable incidents	36241	Pacemaker/defibrillator lead	CT1684	Cardiac pacemakers and associated devices	F1905	Device Revision or Replacement	A0101	Patient-Device Incompatibility		
8785	2019	All other reportable incidents	35852	Single-chamber implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E2114	Perforation	A0101	Patient-Device Incompatibility	29/03/2019	
							F1905	Device Revision or Replacement	A0101	Patient-Device Incompatibility	29/03/2019	
8786	2019	All other reportable incidents	47265	Dual-chamber implantable pacemaker, rate-responsive	CT141	Electrodes/Leads and associated devices	F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A070101	Failure to Capture		
					CT145	Cardiovascular devices	F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A070101	Failure to Capture		
					CT1684	Cardiac pacemakers and associated devices	F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A070101	Failure to Capture		
					CT1685	Cardiac pacemakers	F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A070101	Failure to Capture		
					CT233	Surgical	F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A070101	Failure to Capture		
					CT2406	Implantable	F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A070101	Failure to Capture		
					CT2407	Active implantable	F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A070101	Failure to Capture		
					CT257	Implantable cardiac pacemakers	F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A070101	Failure to Capture		
					CT2727	Active implant devices/ancillaries	F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A070101	Failure to Capture		
					CT2750	Implantable cardiac electrical stimulators and associated devices	F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A070101	Failure to Capture		
					CT2866	Implantable cardiac electrical stimulators	F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A070101	Failure to Capture		
					CT2985	Active	F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A070101	Failure to Capture		
					CT3001	Implantable metal	F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A070101	Failure to Capture		
					CT3002	Implantable synthetic polymer	F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A070101	Failure to Capture		
					CT3006	Implantable inorganic material	F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A070101	Failure to Capture		
					CT314	Electrically powered	F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A070101	Failure to Capture		
					CT321	Long-term surgical invasive	F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A070101	Failure to Capture		
					CT449	Generators and associated devices	F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A070101	Failure to Capture		
					CT981	Single use	F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A070101	Failure to Capture		
					CT983	Surgical invasive	F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A070101	Failure to Capture		
8787	2020	All other reportable incidents	35223	Endocardial/interventricular septal pacing lead	CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070909	Oversensing	21/01/2020	21/01/2020
							F1901	Additional Surgery	A070909	Oversensing	21/01/2020	21/01/2020
							F1905	Device Revision or Replacement	A070909	Oversensing	21/01/2020	21/01/2020
8788	2020	All other reportable incidents	17882	Physiologic-monitoring defibrillation system	CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070903	Failure to Analyze Signal		
							F26	No Health Consequences or Impact	A070903	Failure to Analyze Signal		
8789	2020	All other reportable incidents	37265	Dual-chamber implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	03/02/2020	03/02/2020
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	03/02/2020	03/02/2020
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery	03/02/2020	03/02/2020
8790	2020	All other reportable incidents	47263	Cardiac resynchronization therapy implantable pacemaker	CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070101	Failure to Capture		
									A070909	Oversensing		
									A072201	High Impedance		
							F15	Recognized Device or Procedural Complication	A070101	Failure to Capture		
									A070909	Oversensing		
									A072201	High Impedance		
					CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A070101	Failure to Capture		
									A070909	Oversensing		
									A072201	High Impedance		
							F15	Recognized Device or Procedural Complication	A070101	Failure to Capture		
									A070909	Oversensing		
									A072201	High Impedance		
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A070101	Failure to Capture		
									A070909	Oversensing		
									A072201	High Impedance		
							F15	Recognized Device or Procedural Complication	A070101	Failure to Capture		
									A070909	Oversensing		
									A072201	High Impedance		

					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A070101	Failure to Capture		
									A070909	Oversensing		
									A072201	High Impedance		
							F15	Recognized Device or Procedural Complication	A070101	Failure to Capture		
									A070909	Oversensing		
									A072201	High Impedance		
8791	2020	All other reportable incidents	47267	Single-chamber implantable pacemaker, rate-responsive	CT141	Electrodes/Leads and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A0203	Defective Device	19/02/2020	19/02/2020
									A070504	Premature Discharge of Battery	19/02/2020	19/02/2020
							F08	Hospitalization or Prolonged Hospitalization	A0203	Defective Device	19/02/2020	19/02/2020
									A070504	Premature Discharge of Battery	19/02/2020	19/02/2020
							F12	Serious Injury/ Illness/ Impairment	A0203	Defective Device	19/02/2020	19/02/2020
									A070504	Premature Discharge of Battery	19/02/2020	19/02/2020
							F1903	Device Explantation	A0203	Defective Device	19/02/2020	19/02/2020
									A070504	Premature Discharge of Battery	19/02/2020	19/02/2020
					CT145	Cardiovascular devices	E2403	No Clinical Signs, Symptoms or Conditions	A0203	Defective Device	19/02/2020	19/02/2020
									A070504	Premature Discharge of Battery	19/02/2020	19/02/2020
							F08	Hospitalization or Prolonged Hospitalization	A0203	Defective Device	19/02/2020	19/02/2020
									A070504	Premature Discharge of Battery	19/02/2020	19/02/2020
							F12	Serious Injury/ Illness/ Impairment	A0203	Defective Device	19/02/2020	19/02/2020
									A070504	Premature Discharge of Battery	19/02/2020	19/02/2020
							F1903	Device Explantation	A0203	Defective Device	19/02/2020	19/02/2020
									A070504	Premature Discharge of Battery	19/02/2020	19/02/2020
					CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A0203	Defective Device	19/02/2020	19/02/2020
									A070504	Premature Discharge of Battery	19/02/2020	19/02/2020
							F08	Hospitalization or Prolonged Hospitalization	A0203	Defective Device	19/02/2020	19/02/2020
									A070504	Premature Discharge of Battery	19/02/2020	19/02/2020
							F12	Serious Injury/ Illness/ Impairment	A0203	Defective Device	19/02/2020	19/02/2020
									A070504	Premature Discharge of Battery	19/02/2020	19/02/2020
							F1903	Device Explantation	A0203	Defective Device	19/02/2020	19/02/2020
									A070504	Premature Discharge of Battery	19/02/2020	19/02/2020
					CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A0203	Defective Device	19/02/2020	19/02/2020
									A070504	Premature Discharge of Battery	19/02/2020	19/02/2020
							F08	Hospitalization or Prolonged Hospitalization	A0203	Defective Device	19/02/2020	19/02/2020
									A070504	Premature Discharge of Battery	19/02/2020	19/02/2020
							F12	Serious Injury/ Illness/ Impairment	A0203	Defective Device	19/02/2020	19/02/2020
									A070504	Premature Discharge of Battery	19/02/2020	19/02/2020
							F1903	Device Explantation	A0203	Defective Device	19/02/2020	19/02/2020
									A070504	Premature Discharge of Battery	19/02/2020	19/02/2020
					CT233	Surgical	E2403	No Clinical Signs, Symptoms or Conditions	A0203	Defective Device	19/02/2020	19/02/2020
									A070504	Premature Discharge of Battery	19/02/2020	19/02/2020
							F08	Hospitalization or Prolonged Hospitalization	A0203	Defective Device	19/02/2020	19/02/2020
									A070504	Premature Discharge of Battery	19/02/2020	19/02/2020
							F12	Serious Injury/ Illness/ Impairment	A0203	Defective Device	19/02/2020	19/02/2020
									A070504	Premature Discharge of Battery	19/02/2020	19/02/2020
							F1903	Device Explantation	A0203	Defective Device	19/02/2020	19/02/2020
									A070504	Premature Discharge of Battery	19/02/2020	19/02/2020
					CT2406	Implantable	E2403	No Clinical Signs, Symptoms or Conditions	A0203	Defective Device	19/02/2020	19/02/2020
									A070504	Premature Discharge of Battery	19/02/2020	19/02/2020
							F08	Hospitalization or Prolonged Hospitalization	A0203	Defective Device	19/02/2020	19/02/2020
									A070504	Premature Discharge of Battery	19/02/2020	19/02/2020
							F12	Serious Injury/ Illness/ Impairment	A0203	Defective Device	19/02/2020	19/02/2020
									A070504	Premature Discharge of Battery	19/02/2020	19/02/2020
							F1903	Device Explantation	A0203	Defective Device	19/02/2020	19/02/2020
									A070504	Premature Discharge of Battery	19/02/2020	19/02/2020
					CT2407	Active implantable	E2403	No Clinical Signs, Symptoms or Conditions	A0203	Defective Device	19/02/2020	19/02/2020
									A070504	Premature Discharge of Battery	19/02/2020	19/02/2020
							F08	Hospitalization or Prolonged Hospitalization	A0203	Defective Device	19/02/2020	19/02/2020
									A070504	Premature Discharge of Battery	19/02/2020	19/02/2020
							F12	Serious Injury/ Illness/ Impairment	A0203	Defective Device	19/02/2020	19/02/2020
									A070504	Premature Discharge of Battery	19/02/2020	19/02/2020
							F1903	Device Explantation	A0203	Defective Device	19/02/2020	19/02/2020
									A070504	Premature Discharge of Battery	19/02/2020	19/02/2020
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A0203	Defective Device	19/02/2020	19/02/2020
									A070504	Premature Discharge of Battery	19/02/2020	19/02/2020
							F08	Hospitalization or Prolonged Hospitalization	A0203	Defective Device	19/02/2020	19/02/2020
									A070504	Premature Discharge of Battery	19/02/2020	19/02/2020
							F12	Serious Injury/ Illness/ Impairment	A0203	Defective Device	19/02/2020	19/02/2020
									A070504	Premature Discharge of Battery	19/02/2020	19/02/2020
							F1903	Device Explantation	A0203	Defective Device	19/02/2020	19/02/2020
									A070504	Premature Discharge of Battery	19/02/2020	19/02/2020
					CT2727	Active implant devices/ancillaries	E2403	No Clinical Signs, Symptoms or Conditions	A0203	Defective Device	19/02/2020	19/02/2020
									A070504	Premature Discharge of Battery	19/02/2020	19/02/2020
							F08	Hospitalization or Prolonged Hospitalization	A0203	Defective Device	19/02/2020	19/02/2020
									A070504	Premature Discharge of Battery	19/02/2020	19/02/2020
							F12	Serious Injury/ Illness/ Impairment	A0203	Defective Device	19/02/2020	19/02/2020
									A070504	Premature Discharge of Battery	19/02/2020	19/02/2020
							F1903	Device Explantation	A0203	Defective Device	19/02/2020	19/02/2020
									A070504	Premature Discharge of Battery	19/02/2020	19/02/2020
					CT2750	Implantable cardiac electrical stimulators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A0203	Defective Device	19/02/2020	19/02/2020
									A070504	Premature Discharge of Battery	19/02/2020	19/02/2020
							F08	Hospitalization or Prolonged Hospitalization	A0203	Defective Device	19/02/2020	19/02/2020
									A070504	Premature Discharge of Battery	19/02/2020	19/02/2020
							F12	Serious Injury/ Illness/ Impairment	A0203	Defective Device	19/02/2020	19/02/2020
									A070504	Premature Discharge of Battery	19/02/2020	19/02/2020
							F1903	Device Explantation	A0203	Defective Device	19/02/2020	19/02/2020
									A070504	Premature Discharge of Battery	19/02/2020	19/02/2020
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A0203	Defective Device	19/02/2020	19/02/2020
									A070504	Premature Discharge of Battery	19/02/2020	19/02/2020

							F08	Hospitalization or Prolonged Hospitalization	A0203	Defective Device	19/02/2020	19/02/2020
									A070504	Premature Discharge of Battery	19/02/2020	19/02/2020
							F12	Serious Injury/ Illness/ Impairment	A0203	Defective Device	19/02/2020	19/02/2020
									A070504	Premature Discharge of Battery	19/02/2020	19/02/2020
							F1903	Device Explantation	A0203	Defective Device	19/02/2020	19/02/2020
									A070504	Premature Discharge of Battery	19/02/2020	19/02/2020
					CT2985	Active	E2403	No Clinical Signs, Symptoms or Conditions	A0203	Defective Device	19/02/2020	19/02/2020
									A070504	Premature Discharge of Battery	19/02/2020	19/02/2020
							F08	Hospitalization or Prolonged Hospitalization	A0203	Defective Device	19/02/2020	19/02/2020
									A070504	Premature Discharge of Battery	19/02/2020	19/02/2020
							F12	Serious Injury/ Illness/ Impairment	A0203	Defective Device	19/02/2020	19/02/2020
									A070504	Premature Discharge of Battery	19/02/2020	19/02/2020
							F1903	Device Explantation	A0203	Defective Device	19/02/2020	19/02/2020
									A070504	Premature Discharge of Battery	19/02/2020	19/02/2020
					CT3001	Implantable metal	E2403	No Clinical Signs, Symptoms or Conditions	A0203	Defective Device	19/02/2020	19/02/2020
									A070504	Premature Discharge of Battery	19/02/2020	19/02/2020
							F08	Hospitalization or Prolonged Hospitalization	A0203	Defective Device	19/02/2020	19/02/2020
									A070504	Premature Discharge of Battery	19/02/2020	19/02/2020
							F12	Serious Injury/ Illness/ Impairment	A0203	Defective Device	19/02/2020	19/02/2020
									A070504	Premature Discharge of Battery	19/02/2020	19/02/2020
							F1903	Device Explantation	A0203	Defective Device	19/02/2020	19/02/2020
									A070504	Premature Discharge of Battery	19/02/2020	19/02/2020
					CT3002	Implantable synthetic polymer	E2403	No Clinical Signs, Symptoms or Conditions	A0203	Defective Device	19/02/2020	19/02/2020
									A070504	Premature Discharge of Battery	19/02/2020	19/02/2020
							F08	Hospitalization or Prolonged Hospitalization	A0203	Defective Device	19/02/2020	19/02/2020
									A070504	Premature Discharge of Battery	19/02/2020	19/02/2020
							F12	Serious Injury/ Illness/ Impairment	A0203	Defective Device	19/02/2020	19/02/2020
									A070504	Premature Discharge of Battery	19/02/2020	19/02/2020
							F1903	Device Explantation	A0203	Defective Device	19/02/2020	19/02/2020
									A070504	Premature Discharge of Battery	19/02/2020	19/02/2020
					CT3006	Implantable inorganic material	E2403	No Clinical Signs, Symptoms or Conditions	A0203	Defective Device	19/02/2020	19/02/2020
									A070504	Premature Discharge of Battery	19/02/2020	19/02/2020
							F08	Hospitalization or Prolonged Hospitalization	A0203	Defective Device	19/02/2020	19/02/2020
									A070504	Premature Discharge of Battery	19/02/2020	19/02/2020
							F12	Serious Injury/ Illness/ Impairment	A0203	Defective Device	19/02/2020	19/02/2020
									A070504	Premature Discharge of Battery	19/02/2020	19/02/2020
							F1903	Device Explantation	A0203	Defective Device	19/02/2020	19/02/2020
									A070504	Premature Discharge of Battery	19/02/2020	19/02/2020
					CT314	Electrically powered	E2403	No Clinical Signs, Symptoms or Conditions	A0203	Defective Device	19/02/2020	19/02/2020
									A070504	Premature Discharge of Battery	19/02/2020	19/02/2020
							F08	Hospitalization or Prolonged Hospitalization	A0203	Defective Device	19/02/2020	19/02/2020
									A070504	Premature Discharge of Battery	19/02/2020	19/02/2020
							F12	Serious Injury/ Illness/ Impairment	A0203	Defective Device	19/02/2020	19/02/2020
									A070504	Premature Discharge of Battery	19/02/2020	19/02/2020
							F1903	Device Explantation	A0203	Defective Device	19/02/2020	19/02/2020
									A070504	Premature Discharge of Battery	19/02/2020	19/02/2020
					CT321	Long-term surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A0203	Defective Device	19/02/2020	19/02/2020
									A070504	Premature Discharge of Battery	19/02/2020	19/02/2020
							F08	Hospitalization or Prolonged Hospitalization	A0203	Defective Device	19/02/2020	19/02/2020
									A070504	Premature Discharge of Battery	19/02/2020	19/02/2020
							F12	Serious Injury/ Illness/ Impairment	A0203	Defective Device	19/02/2020	19/02/2020
									A070504	Premature Discharge of Battery	19/02/2020	19/02/2020
							F1903	Device Explantation	A0203	Defective Device	19/02/2020	19/02/2020
									A070504	Premature Discharge of Battery	19/02/2020	19/02/2020
					CT449	Generators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A0203	Defective Device	19/02/2020	19/02/2020
									A070504	Premature Discharge of Battery	19/02/2020	19/02/2020
							F08	Hospitalization or Prolonged Hospitalization	A0203	Defective Device	19/02/2020	19/02/2020
									A070504	Premature Discharge of Battery	19/02/2020	19/02/2020
							F12	Serious Injury/ Illness/ Impairment	A0203	Defective Device	19/02/2020	19/02/2020
									A070504	Premature Discharge of Battery	19/02/2020	19/02/2020
							F1903	Device Explantation	A0203	Defective Device	19/02/2020	19/02/2020
									A070504	Premature Discharge of Battery	19/02/2020	19/02/2020
					CT981	Single use	E2403	No Clinical Signs, Symptoms or Conditions	A0203	Defective Device	19/02/2020	19/02/2020
									A070504	Premature Discharge of Battery	19/02/2020	19/02/2020
							F08	Hospitalization or Prolonged Hospitalization	A0203	Defective Device	19/02/2020	19/02/2020
									A070504	Premature Discharge of Battery	19/02/2020	19/02/2020
							F12	Serious Injury/ Illness/ Impairment	A0203	Defective Device	19/02/2020	19/02/2020
									A070504	Premature Discharge of Battery	19/02/2020	19/02/2020
							F1903	Device Explantation	A0203	Defective Device	19/02/2020	19/02/2020
									A070504	Premature Discharge of Battery	19/02/2020	19/02/2020
					CT983	Surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A0203	Defective Device	19/02/2020	19/02/2020
									A070504	Premature Discharge of Battery	19/02/2020	19/02/2020
							F08	Hospitalization or Prolonged Hospitalization	A0203	Defective Device	19/02/2020	19/02/2020
									A070504	Premature Discharge of Battery	19/02/2020	19/02/2020
							F12	Serious Injury/ Illness/ Impairment	A0203	Defective Device	19/02/2020	19/02/2020
									A070504	Premature Discharge of Battery	19/02/2020	19/02/2020
							F1903	Device Explantation	A0203	Defective Device	19/02/2020	19/02/2020
									A070504	Premature Discharge of Battery	19/02/2020	19/02/2020
8792	2020	All other reportable incidents	17882	Physiologic-monitoring defibrillation system	CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A071301	Failure to Deliver Shock/Stimulation		
							F26	No Health Consequences or Impact	A071301	Failure to Deliver Shock/Stimulation		
8793	2019	All other reportable incidents	47265	Dual-chamber implantable pacemaker, rate-responsive	CT141	Electrodes/Leads and associated devices	F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	17/04/2019	
					CT145	Cardiovascular devices	F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	17/04/2019	
					CT1684	Cardiac pacemakers and associated devices	F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	17/04/2019	
					CT1685	Cardiac pacemakers	F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	17/04/2019	
					CT233	Surgical	F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	17/04/2019	
					CT2406	Implantable	F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	17/04/2019	
					CT2407	Active implantable	F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	17/04/2019	
					CT257	Implantable cardiac pacemakers	F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	17/04/2019	

					CT2727	Active implant devices/ancillaries	F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	17/04/2019		
					CT2750	Implantable cardiac electrical stimulators and associated devices	F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	17/04/2019		
					CT2866	Implantable cardiac electrical stimulators	F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	17/04/2019		
					CT2985	Active	F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	17/04/2019		
					CT3001	Implantable metal	F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	17/04/2019		
					CT3002	Implantable synthetic polymer	F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	17/04/2019		
					CT3006	Implantable inorganic material	F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	17/04/2019		
					CT314	Electrically powered	F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	17/04/2019		
					CT321	Long-term surgical invasive	F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	17/04/2019		
					CT449	Generators and associated devices	F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	17/04/2019		
					CT981	Single use	F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	17/04/2019		
					CT983	Surgical invasive	F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	17/04/2019		
8794	2019	All other reportable incidents	36241	Pacemaker/defibrillator lead	CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070909	Oversensing			
8795	2019	All other reportable incidents	36241	Pacemaker/defibrillator lead	CT1684	Cardiac pacemakers and associated devices	F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A070909	Oversensing			
8796	2019	All other reportable incidents	47265	Dual-chamber implantable pacemaker, rate-responsive	CT141	Electrodes/Leads and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070909	Oversensing			
					CT145	Cardiovascular devices	E2403	No Clinical Signs, Symptoms or Conditions	A070909	Oversensing			
					CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070909	Oversensing			
					CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A070909	Oversensing			
					CT233	Surgical	E2403	No Clinical Signs, Symptoms or Conditions	A070909	Oversensing			
					CT2406	Implantable	E2403	No Clinical Signs, Symptoms or Conditions	A070909	Oversensing			
					CT2407	Active implantable	E2403	No Clinical Signs, Symptoms or Conditions	A070909	Oversensing			
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A070909	Oversensing			
					CT2727	Active implant devices/ancillaries	E2403	No Clinical Signs, Symptoms or Conditions	A070909	Oversensing			
					CT2750	Implantable cardiac electrical stimulators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070909	Oversensing			
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A070909	Oversensing			
					CT2985	Active	E2403	No Clinical Signs, Symptoms or Conditions	A070909	Oversensing			
					CT3001	Implantable metal	E2403	No Clinical Signs, Symptoms or Conditions	A070909	Oversensing			
					CT3002	Implantable synthetic polymer	E2403	No Clinical Signs, Symptoms or Conditions	A070909	Oversensing			
					CT3006	Implantable inorganic material	E2403	No Clinical Signs, Symptoms or Conditions	A070909	Oversensing			
					CT314	Electrically powered	E2403	No Clinical Signs, Symptoms or Conditions	A070909	Oversensing			
					CT321	Long-term surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A070909	Oversensing			
					CT449	Generators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070909	Oversensing			
					CT981	Single use	E2403	No Clinical Signs, Symptoms or Conditions	A070909	Oversensing			
					CT983	Surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A070909	Oversensing			
8797	2019	All other reportable incidents	36241	Pacemaker/defibrillator lead	CT1684	Cardiac pacemakers and associated devices	F14	Prolonged Episode of Care	A010402	Migration			
8798	2020	All other reportable incidents	35223	Endocardial/interventricular septal pacing lead	CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070909	Oversensing			
							F1901	Additional Surgery	A070909	Oversensing			
8799	2020	All other reportable incidents	60789	Intracardiac pacemaker	CT130	Catheters and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A0101	Patient-Device Incompatibility			
									A070504	Premature Discharge of Battery			
									A071102	Failure to Interrogate			
							F1901	Additional Surgery	A0101	Patient-Device Incompatibility			
									A070504	Premature Discharge of Battery			
									A071102	Failure to Interrogate			
					CT145	Cardiovascular devices	E2403	No Clinical Signs, Symptoms or Conditions	A0101	Patient-Device Incompatibility			
									A070504	Premature Discharge of Battery			
									A071102	Failure to Interrogate			
							F1901	Additional Surgery	A0101	Patient-Device Incompatibility			
									A070504	Premature Discharge of Battery			
									A071102	Failure to Interrogate			
					CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A0101	Patient-Device Incompatibility			
									A070504	Premature Discharge of Battery			
									A071102	Failure to Interrogate			
							F1901	Additional Surgery	A0101	Patient-Device Incompatibility			
									A070504	Premature Discharge of Battery			
									A071102	Failure to Interrogate			
					CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A0101	Patient-Device Incompatibility			
									A070504	Premature Discharge of Battery			
									A071102	Failure to Interrogate			
							F1901	Additional Surgery	A0101	Patient-Device Incompatibility			
									A070504	Premature Discharge of Battery			
									A071102	Failure to Interrogate			
					CT233	Surgical	E2403	No Clinical Signs, Symptoms or Conditions	A0101	Patient-Device Incompatibility			
									A070504	Premature Discharge of Battery			
									A071102	Failure to Interrogate			
							F1901	Additional Surgery	A0101	Patient-Device Incompatibility			
									A070504	Premature Discharge of Battery			
									A071102	Failure to Interrogate			
					CT2406	Implantable	E2403	No Clinical Signs, Symptoms or Conditions	A0101	Patient-Device Incompatibility			
									A070504	Premature Discharge of Battery			
									A071102	Failure to Interrogate			
							F1901	Additional Surgery	A0101	Patient-Device Incompatibility			
									A070504	Premature Discharge of Battery			
									A071102	Failure to Interrogate			
					CT2407	Active implantable	E2403	No Clinical Signs, Symptoms or Conditions	A0101	Patient-Device Incompatibility			
									A070504	Premature Discharge of Battery			
									A071102	Failure to Interrogate			
							F1901	Additional Surgery	A0101	Patient-Device Incompatibility			
									A070504	Premature Discharge of Battery			
									A071102	Failure to Interrogate			
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A0101	Patient-Device Incompatibility			
									A070504	Premature Discharge of Battery			
									A071102	Failure to Interrogate			
							F1901	Additional Surgery	A0101	Patient-Device Incompatibility			
									A070504	Premature Discharge of Battery			
									A071102	Failure to Interrogate			
					CT2727	Active implant devices/ancillaries	E2403	No Clinical Signs, Symptoms or Conditions	A0101	Patient-Device Incompatibility			

									A070504	Premature Discharge of Battery		
									A071102	Failure to Interrogate		
							F1901	Additional Surgery	A0101	Patient-Device Incompatibility		
									A070504	Premature Discharge of Battery		
									A071102	Failure to Interrogate		
					CT2750	Implantable cardiac electrical stimulators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A0101	Patient-Device Incompatibility		
									A070504	Premature Discharge of Battery		
									A071102	Failure to Interrogate		
							F1901	Additional Surgery	A0101	Patient-Device Incompatibility		
									A070504	Premature Discharge of Battery		
									A071102	Failure to Interrogate		
					CT2818	Catheterization/Cannulation/Fluid path devices	E2403	No Clinical Signs, Symptoms or Conditions	A0101	Patient-Device Incompatibility		
									A070504	Premature Discharge of Battery		
									A071102	Failure to Interrogate		
							F1901	Additional Surgery	A0101	Patient-Device Incompatibility		
									A070504	Premature Discharge of Battery		
									A071102	Failure to Interrogate		
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A0101	Patient-Device Incompatibility		
									A070504	Premature Discharge of Battery		
									A071102	Failure to Interrogate		
							F1901	Additional Surgery	A0101	Patient-Device Incompatibility		
									A070504	Premature Discharge of Battery		
									A071102	Failure to Interrogate		
					CT2985	Active	E2403	No Clinical Signs, Symptoms or Conditions	A0101	Patient-Device Incompatibility		
									A070504	Premature Discharge of Battery		
									A071102	Failure to Interrogate		
							F1901	Additional Surgery	A0101	Patient-Device Incompatibility		
									A070504	Premature Discharge of Battery		
									A071102	Failure to Interrogate		
					CT3001	Implantable metal	E2403	No Clinical Signs, Symptoms or Conditions	A0101	Patient-Device Incompatibility		
									A070504	Premature Discharge of Battery		
									A071102	Failure to Interrogate		
							F1901	Additional Surgery	A0101	Patient-Device Incompatibility		
									A070504	Premature Discharge of Battery		
									A071102	Failure to Interrogate		
					CT3002	Implantable synthetic polymer	E2403	No Clinical Signs, Symptoms or Conditions	A0101	Patient-Device Incompatibility		
									A070504	Premature Discharge of Battery		
									A071102	Failure to Interrogate		
							F1901	Additional Surgery	A0101	Patient-Device Incompatibility		
									A070504	Premature Discharge of Battery		
									A071102	Failure to Interrogate		
					CT3006	Implantable inorganic material	E2403	No Clinical Signs, Symptoms or Conditions	A0101	Patient-Device Incompatibility		
									A070504	Premature Discharge of Battery		
									A071102	Failure to Interrogate		
							F1901	Additional Surgery	A0101	Patient-Device Incompatibility		
									A070504	Premature Discharge of Battery		
									A071102	Failure to Interrogate		
					CT314	Electrically powered	E2403	No Clinical Signs, Symptoms or Conditions	A0101	Patient-Device Incompatibility		
									A070504	Premature Discharge of Battery		
									A071102	Failure to Interrogate		
							F1901	Additional Surgery	A0101	Patient-Device Incompatibility		
									A070504	Premature Discharge of Battery		
									A071102	Failure to Interrogate		
					CT321	Long-term surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A0101	Patient-Device Incompatibility		
									A070504	Premature Discharge of Battery		
									A071102	Failure to Interrogate		
							F1901	Additional Surgery	A0101	Patient-Device Incompatibility		
									A070504	Premature Discharge of Battery		
									A071102	Failure to Interrogate		
					CT981	Single use	E2403	No Clinical Signs, Symptoms or Conditions	A0101	Patient-Device Incompatibility		
									A070504	Premature Discharge of Battery		
									A071102	Failure to Interrogate		
							F1901	Additional Surgery	A0101	Patient-Device Incompatibility		
									A070504	Premature Discharge of Battery		
									A071102	Failure to Interrogate		
					CT983	Surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A0101	Patient-Device Incompatibility		
									A070504	Premature Discharge of Battery		
									A071102	Failure to Interrogate		
							F1901	Additional Surgery	A0101	Patient-Device Incompatibility		
									A070504	Premature Discharge of Battery		
									A071102	Failure to Interrogate		
8800	2020	All other reportable incidents	37265	Dual-chamber implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	24/12/2019	24/12/2019
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery	24/12/2019	24/12/2019
8801	2020	All other reportable incidents	35223	Endocardial/interventricular septal pacing lead	CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070909	Oversensing		
									A0722	Impedance Problem		
									A072201	High Impedance		
									A090801	Signal Artifact/Noise		
							F26	No Health Consequences or Impact	A070909	Oversensing		
									A0722	Impedance Problem		
									A072201	High Impedance		
									A090801	Signal Artifact/Noise		
8802	2020	All other reportable incidents	36241	Pacemaker/defibrillator lead	CT1684	Cardiac pacemakers and associated devices	F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A070101	Failure to Capture		
									A070909	Oversensing		
8803	2020	All other reportable incidents	47267	Single-chamber implantable pacemaker, rate-responsive	CT141	Electrodes/Leads and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A0908	Incorrect, Inadequate or Imprecise Result or Readings	28/01/2020	28/01/2020
					CT145	Cardiovascular devices	E2403	No Clinical Signs, Symptoms or Conditions	A0908	Incorrect, Inadequate or Imprecise Result or Readings	28/01/2020	28/01/2020

					CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A0908	Incorrect, Inadequate or Imprecise Result or Readings	28/01/2020	28/01/2020
					CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A0908	Incorrect, Inadequate or Imprecise Result or Readings	28/01/2020	28/01/2020
					CT233	Surgical	E2403	No Clinical Signs, Symptoms or Conditions	A0908	Incorrect, Inadequate or Imprecise Result or Readings	28/01/2020	28/01/2020
					CT2406	Implantable	E2403	No Clinical Signs, Symptoms or Conditions	A0908	Incorrect, Inadequate or Imprecise Result or Readings	28/01/2020	28/01/2020
					CT2407	Active implantable	E2403	No Clinical Signs, Symptoms or Conditions	A0908	Incorrect, Inadequate or Imprecise Result or Readings	28/01/2020	28/01/2020
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A0908	Incorrect, Inadequate or Imprecise Result or Readings	28/01/2020	28/01/2020
					CT2727	Active implant devices/ancillaries	E2403	No Clinical Signs, Symptoms or Conditions	A0908	Incorrect, Inadequate or Imprecise Result or Readings	28/01/2020	28/01/2020
					CT2750	Implantable cardiac electrical stimulators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A0908	Incorrect, Inadequate or Imprecise Result or Readings	28/01/2020	28/01/2020
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A0908	Incorrect, Inadequate or Imprecise Result or Readings	28/01/2020	28/01/2020
					CT2985	Active	E2403	No Clinical Signs, Symptoms or Conditions	A0908	Incorrect, Inadequate or Imprecise Result or Readings	28/01/2020	28/01/2020
					CT3001	Implantable metal	E2403	No Clinical Signs, Symptoms or Conditions	A0908	Incorrect, Inadequate or Imprecise Result or Readings	28/01/2020	28/01/2020
					CT3002	Implantable synthetic polymer	E2403	No Clinical Signs, Symptoms or Conditions	A0908	Incorrect, Inadequate or Imprecise Result or Readings	28/01/2020	28/01/2020
					CT3006	Implantable inorganic material	E2403	No Clinical Signs, Symptoms or Conditions	A0908	Incorrect, Inadequate or Imprecise Result or Readings	28/01/2020	28/01/2020
					CT314	Electrically powered	E2403	No Clinical Signs, Symptoms or Conditions	A0908	Incorrect, Inadequate or Imprecise Result or Readings	28/01/2020	28/01/2020
					CT321	Long-term surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A0908	Incorrect, Inadequate or Imprecise Result or Readings	28/01/2020	28/01/2020
					CT449	Generators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A0908	Incorrect, Inadequate or Imprecise Result or Readings	28/01/2020	28/01/2020
					CT981	Single use	E2403	No Clinical Signs, Symptoms or Conditions	A0908	Incorrect, Inadequate or Imprecise Result or Readings	28/01/2020	28/01/2020
					CT983	Surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A0908	Incorrect, Inadequate or Imprecise Result or Readings	28/01/2020	28/01/2020
8804	2019	All other reportable incidents	47265	Dual-chamber implantable pacemaker, rate-responsive	CT141	Electrodes/Leads and associated devices	F1905	Device Revision or Replacement	A071206	Pacing Asynchronously	21/03/2019	
					CT145	Cardiovascular devices	F1905	Device Revision or Replacement	A071206	Pacing Asynchronously	21/03/2019	
					CT1684	Cardiac pacemakers and associated devices	F1905	Device Revision or Replacement	A071206	Pacing Asynchronously	21/03/2019	
					CT1685	Cardiac pacemakers	F1905	Device Revision or Replacement	A071206	Pacing Asynchronously	21/03/2019	
					CT233	Surgical	F1905	Device Revision or Replacement	A071206	Pacing Asynchronously	21/03/2019	
					CT2406	Implantable	F1905	Device Revision or Replacement	A071206	Pacing Asynchronously	21/03/2019	
					CT2407	Active implantable	F1905	Device Revision or Replacement	A071206	Pacing Asynchronously	21/03/2019	
					CT257	Implantable cardiac pacemakers	F1905	Device Revision or Replacement	A071206	Pacing Asynchronously	21/03/2019	
					CT2727	Active implant devices/ancillaries	F1905	Device Revision or Replacement	A071206	Pacing Asynchronously	21/03/2019	
					CT2750	Implantable cardiac electrical stimulators and associated devices	F1905	Device Revision or Replacement	A071206	Pacing Asynchronously	21/03/2019	
					CT2866	Implantable cardiac electrical stimulators	F1905	Device Revision or Replacement	A071206	Pacing Asynchronously	21/03/2019	
					CT2985	Active	F1905	Device Revision or Replacement	A071206	Pacing Asynchronously	21/03/2019	
					CT3001	Implantable metal	F1905	Device Revision or Replacement	A071206	Pacing Asynchronously	21/03/2019	
					CT3002	Implantable synthetic polymer	F1905	Device Revision or Replacement	A071206	Pacing Asynchronously	21/03/2019	
					CT3006	Implantable inorganic material	F1905	Device Revision or Replacement	A071206	Pacing Asynchronously	21/03/2019	
					CT314	Electrically powered	F1905	Device Revision or Replacement	A071206	Pacing Asynchronously	21/03/2019	
					CT321	Long-term surgical invasive	F1905	Device Revision or Replacement	A071206	Pacing Asynchronously	21/03/2019	
					CT449	Generators and associated devices	F1905	Device Revision or Replacement	A071206	Pacing Asynchronously	21/03/2019	
					CT981	Single use	F1905	Device Revision or Replacement	A071206	Pacing Asynchronously	21/03/2019	
					CT983	Surgical invasive	F1905	Device Revision or Replacement	A071206	Pacing Asynchronously	21/03/2019	
8805	2020	All other reportable incidents	35852	Single-chamber implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A070909	Oversensing		
									A071302	Inappropriate/Inadequate Shock/Stimulation		
8806	2020	All other reportable incidents	35852	Single-chamber implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A070909	Oversensing		
									A071302	Inappropriate/Inadequate Shock/Stimulation		
8807	2019	Death	17882	Physiologic-monitoring defibrillation system	CT1684	Cardiac pacemakers and associated devices	F05	Delay to Treatment/ Therapy	A070908	Failure to Sense		
8808	2020	All other reportable incidents	37265	Dual-chamber implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	13/02/2020	13/02/2020
							F19	Surgical Intervention	A070504	Premature Discharge of Battery	13/02/2020	13/02/2020
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	13/02/2020	13/02/2020
8809	2020	All other reportable incidents	36241	Pacemaker/defibrillator lead	CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A0722	Impedance Problem	29/01/2020	29/01/2020
8810	2020	All other reportable incidents	47265	Dual-chamber implantable pacemaker, rate-responsive	CT141	Electrodes/Leads and associated devices	E2326	Inflammation	A0908	Incorrect, Inadequate or Imprecise Result or Readings	13/02/2020	13/02/2020
							E2330	Pain	A0908	Incorrect, Inadequate or Imprecise Result or Readings	13/02/2020	13/02/2020
							E2338	Swelling/ Edema	A0908	Incorrect, Inadequate or Imprecise Result or Readings	13/02/2020	13/02/2020
							F08	Hospitalization or Prolonged Hospitalization	A0908	Incorrect, Inadequate or Imprecise Result or Readings	13/02/2020	13/02/2020
							F12	Serious Injury/ Illness/ Impairment	A0908	Incorrect, Inadequate or Imprecise Result or Readings	13/02/2020	13/02/2020
					CT145	Cardiovascular devices	E2326	Inflammation	A0908	Incorrect, Inadequate or Imprecise Result or Readings	13/02/2020	13/02/2020
							E2330	Pain	A0908	Incorrect, Inadequate or Imprecise Result or Readings	13/02/2020	13/02/2020
							E2338	Swelling/ Edema	A0908	Incorrect, Inadequate or Imprecise Result or Readings	13/02/2020	13/02/2020
							F08	Hospitalization or Prolonged Hospitalization	A0908	Incorrect, Inadequate or Imprecise Result or Readings	13/02/2020	13/02/2020
							F12	Serious Injury/ Illness/ Impairment	A0908	Incorrect, Inadequate or Imprecise Result or Readings	13/02/2020	13/02/2020
					CT1684	Cardiac pacemakers and associated devices	E2326	Inflammation	A0908	Incorrect, Inadequate or Imprecise Result or Readings	13/02/2020	13/02/2020
							E2330	Pain	A0908	Incorrect, Inadequate or Imprecise Result or Readings	13/02/2020	13/02/2020
							E2338	Swelling/ Edema	A0908	Incorrect, Inadequate or Imprecise Result or Readings	13/02/2020	13/02/2020
							F08	Hospitalization or Prolonged Hospitalization	A0908	Incorrect, Inadequate or Imprecise Result or Readings	13/02/2020	13/02/2020
							F12	Serious Injury/ Illness/ Impairment	A0908	Incorrect, Inadequate or Imprecise Result or Readings	13/02/2020	13/02/2020

					CT1685	Cardiac pacemakers	E2326	Inflammation	A0908	Incorrect, Inadequate or Imprecise Result or Readings	13/02/2020	13/02/2020
							E2330	Pain	A0908	Incorrect, Inadequate or Imprecise Result or Readings	13/02/2020	13/02/2020
							E2338	Swelling/ Edema	A0908	Incorrect, Inadequate or Imprecise Result or Readings	13/02/2020	13/02/2020
							F08	Hospitalization or Prolonged Hospitalization	A0908	Incorrect, Inadequate or Imprecise Result or Readings	13/02/2020	13/02/2020
							F12	Serious Injury/ Illness/ Impairment	A0908	Incorrect, Inadequate or Imprecise Result or Readings	13/02/2020	13/02/2020
					CT233	Surgical	E2326	Inflammation	A0908	Incorrect, Inadequate or Imprecise Result or Readings	13/02/2020	13/02/2020
							E2330	Pain	A0908	Incorrect, Inadequate or Imprecise Result or Readings	13/02/2020	13/02/2020
							E2338	Swelling/ Edema	A0908	Incorrect, Inadequate or Imprecise Result or Readings	13/02/2020	13/02/2020
							F08	Hospitalization or Prolonged Hospitalization	A0908	Incorrect, Inadequate or Imprecise Result or Readings	13/02/2020	13/02/2020
							F12	Serious Injury/ Illness/ Impairment	A0908	Incorrect, Inadequate or Imprecise Result or Readings	13/02/2020	13/02/2020
					CT2406	Implantable	E2326	Inflammation	A0908	Incorrect, Inadequate or Imprecise Result or Readings	13/02/2020	13/02/2020
							E2330	Pain	A0908	Incorrect, Inadequate or Imprecise Result or Readings	13/02/2020	13/02/2020
							E2338	Swelling/ Edema	A0908	Incorrect, Inadequate or Imprecise Result or Readings	13/02/2020	13/02/2020
							F08	Hospitalization or Prolonged Hospitalization	A0908	Incorrect, Inadequate or Imprecise Result or Readings	13/02/2020	13/02/2020
							F12	Serious Injury/ Illness/ Impairment	A0908	Incorrect, Inadequate or Imprecise Result or Readings	13/02/2020	13/02/2020
					CT2407	Active implantable	E2326	Inflammation	A0908	Incorrect, Inadequate or Imprecise Result or Readings	13/02/2020	13/02/2020
							E2330	Pain	A0908	Incorrect, Inadequate or Imprecise Result or Readings	13/02/2020	13/02/2020
							E2338	Swelling/ Edema	A0908	Incorrect, Inadequate or Imprecise Result or Readings	13/02/2020	13/02/2020
							F08	Hospitalization or Prolonged Hospitalization	A0908	Incorrect, Inadequate or Imprecise Result or Readings	13/02/2020	13/02/2020
							F12	Serious Injury/ Illness/ Impairment	A0908	Incorrect, Inadequate or Imprecise Result or Readings	13/02/2020	13/02/2020
					CT257	Implantable cardiac pacemakers	E2326	Inflammation	A0908	Incorrect, Inadequate or Imprecise Result or Readings	13/02/2020	13/02/2020
							E2330	Pain	A0908	Incorrect, Inadequate or Imprecise Result or Readings	13/02/2020	13/02/2020
							E2338	Swelling/ Edema	A0908	Incorrect, Inadequate or Imprecise Result or Readings	13/02/2020	13/02/2020
							F08	Hospitalization or Prolonged Hospitalization	A0908	Incorrect, Inadequate or Imprecise Result or Readings	13/02/2020	13/02/2020
							F12	Serious Injury/ Illness/ Impairment	A0908	Incorrect, Inadequate or Imprecise Result or Readings	13/02/2020	13/02/2020
					CT2727	Active implant devices/ancillaries	E2326	Inflammation	A0908	Incorrect, Inadequate or Imprecise Result or Readings	13/02/2020	13/02/2020
							E2330	Pain	A0908	Incorrect, Inadequate or Imprecise Result or Readings	13/02/2020	13/02/2020
							E2338	Swelling/ Edema	A0908	Incorrect, Inadequate or Imprecise Result or Readings	13/02/2020	13/02/2020
							F08	Hospitalization or Prolonged Hospitalization	A0908	Incorrect, Inadequate or Imprecise Result or Readings	13/02/2020	13/02/2020
							F12	Serious Injury/ Illness/ Impairment	A0908	Incorrect, Inadequate or Imprecise Result or Readings	13/02/2020	13/02/2020
					CT2750	Implantable cardiac electrical stimulators and associated devices	E2326	Inflammation	A0908	Incorrect, Inadequate or Imprecise Result or Readings	13/02/2020	13/02/2020
							E2330	Pain	A0908	Incorrect, Inadequate or Imprecise Result or Readings	13/02/2020	13/02/2020
							E2338	Swelling/ Edema	A0908	Incorrect, Inadequate or Imprecise Result or Readings	13/02/2020	13/02/2020
							F08	Hospitalization or Prolonged Hospitalization	A0908	Incorrect, Inadequate or Imprecise Result or Readings	13/02/2020	13/02/2020
							F12	Serious Injury/ Illness/ Impairment	A0908	Incorrect, Inadequate or Imprecise Result or Readings	13/02/2020	13/02/2020
					CT2866	Implantable cardiac electrical stimulators	E2326	Inflammation	A0908	Incorrect, Inadequate or Imprecise Result or Readings	13/02/2020	13/02/2020
							E2330	Pain	A0908	Incorrect, Inadequate or Imprecise Result or Readings	13/02/2020	13/02/2020
							E2338	Swelling/ Edema	A0908	Incorrect, Inadequate or Imprecise Result or Readings	13/02/2020	13/02/2020
							F08	Hospitalization or Prolonged Hospitalization	A0908	Incorrect, Inadequate or Imprecise Result or Readings	13/02/2020	13/02/2020
							F12	Serious Injury/ Illness/ Impairment	A0908	Incorrect, Inadequate or Imprecise Result or Readings	13/02/2020	13/02/2020
					CT2985	Active	E2326	Inflammation	A0908	Incorrect, Inadequate or Imprecise Result or Readings	13/02/2020	13/02/2020
							E2330	Pain	A0908	Incorrect, Inadequate or Imprecise Result or Readings	13/02/2020	13/02/2020
							E2338	Swelling/ Edema	A0908	Incorrect, Inadequate or Imprecise Result or Readings	13/02/2020	13/02/2020
							F08	Hospitalization or Prolonged Hospitalization	A0908	Incorrect, Inadequate or Imprecise Result or Readings	13/02/2020	13/02/2020
							F12	Serious Injury/ Illness/ Impairment	A0908	Incorrect, Inadequate or Imprecise Result or Readings	13/02/2020	13/02/2020
					CT3001	Implantable metal	E2326	Inflammation	A0908	Incorrect, Inadequate or Imprecise Result or Readings	13/02/2020	13/02/2020
							E2330	Pain	A0908	Incorrect, Inadequate or Imprecise Result or Readings	13/02/2020	13/02/2020
							E2338	Swelling/ Edema	A0908	Incorrect, Inadequate or Imprecise Result or Readings	13/02/2020	13/02/2020
							F08	Hospitalization or Prolonged Hospitalization	A0908	Incorrect, Inadequate or Imprecise Result or Readings	13/02/2020	13/02/2020
							F12	Serious Injury/ Illness/ Impairment	A0908	Incorrect, Inadequate or Imprecise Result or Readings	13/02/2020	13/02/2020

					CT3002	Implantable synthetic polymer	E2326	Inflammation	A0908	Incorrect, Inadequate or Imprecise Result or Readings	13/02/2020	13/02/2020
							E2330	Pain	A0908	Incorrect, Inadequate or Imprecise Result or Readings	13/02/2020	13/02/2020
							E2338	Swelling/ Edema	A0908	Incorrect, Inadequate or Imprecise Result or Readings	13/02/2020	13/02/2020
							F08	Hospitalization or Prolonged Hospitalization	A0908	Incorrect, Inadequate or Imprecise Result or Readings	13/02/2020	13/02/2020
							F12	Serious Injury/ Illness/ Impairment	A0908	Incorrect, Inadequate or Imprecise Result or Readings	13/02/2020	13/02/2020
					CT3006	Implantable inorganic material	E2326	Inflammation	A0908	Incorrect, Inadequate or Imprecise Result or Readings	13/02/2020	13/02/2020
							E2330	Pain	A0908	Incorrect, Inadequate or Imprecise Result or Readings	13/02/2020	13/02/2020
							E2338	Swelling/ Edema	A0908	Incorrect, Inadequate or Imprecise Result or Readings	13/02/2020	13/02/2020
							F08	Hospitalization or Prolonged Hospitalization	A0908	Incorrect, Inadequate or Imprecise Result or Readings	13/02/2020	13/02/2020
							F12	Serious Injury/ Illness/ Impairment	A0908	Incorrect, Inadequate or Imprecise Result or Readings	13/02/2020	13/02/2020
					CT314	Electrically powered	E2326	Inflammation	A0908	Incorrect, Inadequate or Imprecise Result or Readings	13/02/2020	13/02/2020
							E2330	Pain	A0908	Incorrect, Inadequate or Imprecise Result or Readings	13/02/2020	13/02/2020
							E2338	Swelling/ Edema	A0908	Incorrect, Inadequate or Imprecise Result or Readings	13/02/2020	13/02/2020
							F08	Hospitalization or Prolonged Hospitalization	A0908	Incorrect, Inadequate or Imprecise Result or Readings	13/02/2020	13/02/2020
							F12	Serious Injury/ Illness/ Impairment	A0908	Incorrect, Inadequate or Imprecise Result or Readings	13/02/2020	13/02/2020
					CT321	Long-term surgical invasive	E2326	Inflammation	A0908	Incorrect, Inadequate or Imprecise Result or Readings	13/02/2020	13/02/2020
							E2330	Pain	A0908	Incorrect, Inadequate or Imprecise Result or Readings	13/02/2020	13/02/2020
							E2338	Swelling/ Edema	A0908	Incorrect, Inadequate or Imprecise Result or Readings	13/02/2020	13/02/2020
							F08	Hospitalization or Prolonged Hospitalization	A0908	Incorrect, Inadequate or Imprecise Result or Readings	13/02/2020	13/02/2020
							F12	Serious Injury/ Illness/ Impairment	A0908	Incorrect, Inadequate or Imprecise Result or Readings	13/02/2020	13/02/2020
					CT449	Generators and associated devices	E2326	Inflammation	A0908	Incorrect, Inadequate or Imprecise Result or Readings	13/02/2020	13/02/2020
							E2330	Pain	A0908	Incorrect, Inadequate or Imprecise Result or Readings	13/02/2020	13/02/2020
							E2338	Swelling/ Edema	A0908	Incorrect, Inadequate or Imprecise Result or Readings	13/02/2020	13/02/2020
							F08	Hospitalization or Prolonged Hospitalization	A0908	Incorrect, Inadequate or Imprecise Result or Readings	13/02/2020	13/02/2020
							F12	Serious Injury/ Illness/ Impairment	A0908	Incorrect, Inadequate or Imprecise Result or Readings	13/02/2020	13/02/2020
					CT981	Single use	E2326	Inflammation	A0908	Incorrect, Inadequate or Imprecise Result or Readings	13/02/2020	13/02/2020
							E2330	Pain	A0908	Incorrect, Inadequate or Imprecise Result or Readings	13/02/2020	13/02/2020
							E2338	Swelling/ Edema	A0908	Incorrect, Inadequate or Imprecise Result or Readings	13/02/2020	13/02/2020
							F08	Hospitalization or Prolonged Hospitalization	A0908	Incorrect, Inadequate or Imprecise Result or Readings	13/02/2020	13/02/2020
							F12	Serious Injury/ Illness/ Impairment	A0908	Incorrect, Inadequate or Imprecise Result or Readings	13/02/2020	13/02/2020
					CT983	Surgical invasive	E2326	Inflammation	A0908	Incorrect, Inadequate or Imprecise Result or Readings	13/02/2020	13/02/2020
							E2330	Pain	A0908	Incorrect, Inadequate or Imprecise Result or Readings	13/02/2020	13/02/2020
							E2338	Swelling/ Edema	A0908	Incorrect, Inadequate or Imprecise Result or Readings	13/02/2020	13/02/2020
							F08	Hospitalization or Prolonged Hospitalization	A0908	Incorrect, Inadequate or Imprecise Result or Readings	13/02/2020	13/02/2020
							F12	Serious Injury/ Illness/ Impairment	A0908	Incorrect, Inadequate or Imprecise Result or Readings	13/02/2020	13/02/2020
8811	2019	All other reportable incidents	60789	Intracardiac pacemaker	CT130	Catheters and associated devices	F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A0712	Pacing Problem		
							F1905	Device Revision or Replacement	A0712	Pacing Problem		
					CT145	Cardiovascular devices	F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A0712	Pacing Problem		
							F1905	Device Revision or Replacement	A0712	Pacing Problem		
					CT1684	Cardiac pacemakers and associated devices	F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A0712	Pacing Problem		
							F1905	Device Revision or Replacement	A0712	Pacing Problem		
					CT1685	Cardiac pacemakers	F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A0712	Pacing Problem		
							F1905	Device Revision or Replacement	A0712	Pacing Problem		
					CT233	Surgical	F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A0712	Pacing Problem		
							F1905	Device Revision or Replacement	A0712	Pacing Problem		
					CT2406	Implantable	F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A0712	Pacing Problem		
							F1905	Device Revision or Replacement	A0712	Pacing Problem		
					CT2407	Active implantable	F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A0712	Pacing Problem		
							F1905	Device Revision or Replacement	A0712	Pacing Problem		
					CT257	Implantable cardiac pacemakers	F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A0712	Pacing Problem		
							F1905	Device Revision or Replacement	A0712	Pacing Problem		
					CT2727	Active implant devices/ancillaries	F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A0712	Pacing Problem		
							F1905	Device Revision or Replacement	A0712	Pacing Problem		
					CT2750	Implantable cardiac electrical stimulators and associated devices	F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A0712	Pacing Problem		
							F1905	Device Revision or Replacement	A0712	Pacing Problem		
					CT2818	Catheterization/Cannulation/Fluid path devices	F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A0712	Pacing Problem		
							F1905	Device Revision or Replacement	A0712	Pacing Problem		
					CT2866	Implantable cardiac electrical stimulators	F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A0712	Pacing Problem		
							F1905	Device Revision or Replacement	A0712	Pacing Problem		
					CT2985	Active	F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A0712	Pacing Problem		
							F1905	Device Revision or Replacement	A0712	Pacing Problem		

					CT3001	Implantable metal	F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A0712	Pacing Problem		
							F1905	Device Revision or Replacement	A0712	Pacing Problem		
					CT3002	Implantable synthetic polymer	F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A0712	Pacing Problem		
							F1905	Device Revision or Replacement	A0712	Pacing Problem		
					CT3006	Implantable inorganic material	F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A0712	Pacing Problem		
							F1905	Device Revision or Replacement	A0712	Pacing Problem		
					CT314	Electrically powered	F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A0712	Pacing Problem		
							F1905	Device Revision or Replacement	A0712	Pacing Problem		
					CT321	Long-term surgical invasive	F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A0712	Pacing Problem		
							F1905	Device Revision or Replacement	A0712	Pacing Problem		
					CT981	Single use	F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A0712	Pacing Problem		
							F1905	Device Revision or Replacement	A0712	Pacing Problem		
					CT983	Surgical invasive	F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A0712	Pacing Problem		
							F1905	Device Revision or Replacement	A0712	Pacing Problem		
8812	2019	All other reportable incidents	36241	Pacemaker/defibrillator lead	CT1684	Cardiac pacemakers and associated devices	F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A072201	High Impedance		
8813	2019	All other reportable incidents	35852	Single-chamber implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A072201	High Impedance		
8814	2020	All other reportable incidents	47267	Single-chamber implantable pacemaker, rate-responsive	CT141	Electrodes/Leads and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A07	Electrical /Electronic Property Problem		
									A090812	Missing Test Results		
					CT145	Cardiovascular devices	E2403	No Clinical Signs, Symptoms or Conditions	A07	Electrical /Electronic Property Problem		
									A090812	Missing Test Results		
					CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A07	Electrical /Electronic Property Problem		
									A090812	Missing Test Results		
					CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A07	Electrical /Electronic Property Problem		
									A090812	Missing Test Results		
					CT233	Surgical	E2403	No Clinical Signs, Symptoms or Conditions	A07	Electrical /Electronic Property Problem		
									A090812	Missing Test Results		
					CT2406	Implantable	E2403	No Clinical Signs, Symptoms or Conditions	A07	Electrical /Electronic Property Problem		
									A090812	Missing Test Results		
					CT2407	Active implantable	E2403	No Clinical Signs, Symptoms or Conditions	A07	Electrical /Electronic Property Problem		
									A090812	Missing Test Results		
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A07	Electrical /Electronic Property Problem		
									A090812	Missing Test Results		
					CT2727	Active implant devices/ancillaries	E2403	No Clinical Signs, Symptoms or Conditions	A07	Electrical /Electronic Property Problem		
									A090812	Missing Test Results		
					CT2750	Implantable cardiac electrical stimulators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A07	Electrical /Electronic Property Problem		
									A090812	Missing Test Results		
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A07	Electrical /Electronic Property Problem		
									A090812	Missing Test Results		
					CT2985	Active	E2403	No Clinical Signs, Symptoms or Conditions	A07	Electrical /Electronic Property Problem		
									A090812	Missing Test Results		
					CT3001	Implantable metal	E2403	No Clinical Signs, Symptoms or Conditions	A07	Electrical /Electronic Property Problem		
									A090812	Missing Test Results		
					CT3002	Implantable synthetic polymer	E2403	No Clinical Signs, Symptoms or Conditions	A07	Electrical /Electronic Property Problem		
									A090812	Missing Test Results		
					CT3006	Implantable inorganic material	E2403	No Clinical Signs, Symptoms or Conditions	A07	Electrical /Electronic Property Problem		
									A090812	Missing Test Results		
					CT314	Electrically powered	E2403	No Clinical Signs, Symptoms or Conditions	A07	Electrical /Electronic Property Problem		
									A090812	Missing Test Results		
					CT321	Long-term surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A07	Electrical /Electronic Property Problem		
									A090812	Missing Test Results		
					CT449	Generators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A07	Electrical /Electronic Property Problem		
									A090812	Missing Test Results		
					CT981	Single use	E2403	No Clinical Signs, Symptoms or Conditions	A07	Electrical /Electronic Property Problem		
									A090812	Missing Test Results		
					CT983	Surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A07	Electrical /Electronic Property Problem		
									A090812	Missing Test Results		
8815	2019	All other reportable incidents	36241	Pacemaker/defibrillator lead	CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A0510	Retraction Problem	28/03/2019	
8816	2019	All other reportable incidents	47270	Cardiac resynchronizatlon therapy implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A071302	Inappropriate/Inadequate Shock/Stimulation		
8817	2019	All other reportable incidents	36241	Pacemaker/defibrillator lead	CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A072201	High Impedance		
8818	2020	All other reportable incidents	60910	Coronary venous pacing lead	CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A040101	Fracture		
									A0501	Detachment of Device or Device Component		
									A070102	High Capture Threshold		
									A072201	High Impedance		
							F08	Hospitalization or Prolonged Hospitalization	A040101	Fracture		
									A0501	Detachment of Device or Device Component		
									A070102	High Capture Threshold		
									A072201	High Impedance		
							F12	Serious Injury/ Illness/ Impairment	A040101	Fracture		
									A0501	Detachment of Device or Device Component		
									A070102	High Capture Threshold		
									A072201	High Impedance		
							F1903	Device Explantation	A040101	Fracture		
									A0501	Detachment of Device or Device Component		
									A070102	High Capture Threshold		
									A072201	High Impedance		
							F23	Unexpected Medical Intervention	A040101	Fracture		
									A0501	Detachment of Device or Device Component		
									A070102	High Capture Threshold		
									A072201	High Impedance		
8819	2020	All other reportable incidents	47265	Dual-chamber implantable pacemaker, rate-responsive	CT141	Electrodes/Leads and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
					CT145	Cardiovascular devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
					CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
					CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
					CT233	Surgical	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
					CT2406	Implantable	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
					CT2407	Active implantable	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		

					CT2727	Active implant devices/ancillaries	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
					CT2750	Implantable cardiac electrical stimulators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
					CT2985	Active	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
					CT3001	Implantable metal	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
					CT3002	Implantable synthetic polymer	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
					CT3006	Implantable inorganic material	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
					CT314	Electrically powered	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
					CT321	Long-term surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
					CT449	Generators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
					CT981	Single use	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
					CT983	Surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
8820	2020	All other reportable incidents	37265	Dual-chamber implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	15/02/2020	15/02/2020
							F19	Surgical Intervention	A070504	Premature Discharge of Battery	15/02/2020	15/02/2020
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	15/02/2020	15/02/2020
8821	2020	All other reportable incidents	37265	Dual-chamber implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	01/03/2020	01/03/2020
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	01/03/2020	01/03/2020
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery	01/03/2020	01/03/2020
8822	2020	All other reportable incidents	37265	Dual-chamber implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E060109	Tachycardia	A070909	Oversensing		
							E2403	No Clinical Signs, Symptoms or Conditions	A070909	Oversensing		
							F26	No Health Consequences or Impact	A070909	Oversensing		
8823	2020	All other reportable incidents	37265	Dual-chamber implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A070909	Oversensing		
							F26	No Health Consequences or Impact	A070909	Oversensing		
8824	2019	All other reportable incidents	37265	Dual-chamber implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A071207	Pacing Inadequately		
8825	2019	All other reportable incidents	47267	Single-chamber implantable pacemaker, rate-responsive	CT141	Electrodes/Leads and associated devices	F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A070504	Premature Discharge of Battery	03/04/2019	
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	03/04/2019	
					CT145	Cardiovascular devices	F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A070504	Premature Discharge of Battery	03/04/2019	
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	03/04/2019	
					CT1684	Cardiac pacemakers and associated devices	F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A070504	Premature Discharge of Battery	03/04/2019	
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	03/04/2019	
					CT1685	Cardiac pacemakers	F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A070504	Premature Discharge of Battery	03/04/2019	
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	03/04/2019	
					CT233	Surgical	F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A070504	Premature Discharge of Battery	03/04/2019	
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	03/04/2019	
					CT2406	Implantable	F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A070504	Premature Discharge of Battery	03/04/2019	
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	03/04/2019	
					CT2407	Active implantable	F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A070504	Premature Discharge of Battery	03/04/2019	
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	03/04/2019	
					CT257	Implantable cardiac pacemakers	F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A070504	Premature Discharge of Battery	03/04/2019	
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	03/04/2019	
					CT2727	Active implant devices/ancillaries	F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A070504	Premature Discharge of Battery	03/04/2019	
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	03/04/2019	
					CT2750	Implantable cardiac electrical stimulators and associated devices	F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A070504	Premature Discharge of Battery	03/04/2019	
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	03/04/2019	
					CT2866	Implantable cardiac electrical stimulators	F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A070504	Premature Discharge of Battery	03/04/2019	
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	03/04/2019	
					CT2985	Active	F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A070504	Premature Discharge of Battery	03/04/2019	
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	03/04/2019	
					CT3001	Implantable metal	F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A070504	Premature Discharge of Battery	03/04/2019	
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	03/04/2019	
					CT3002	Implantable synthetic polymer	F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A070504	Premature Discharge of Battery	03/04/2019	
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	03/04/2019	
					CT3006	Implantable inorganic material	F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A070504	Premature Discharge of Battery	03/04/2019	
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	03/04/2019	
					CT314	Electrically powered	F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A070504	Premature Discharge of Battery	03/04/2019	
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	03/04/2019	
					CT321	Long-term surgical invasive	F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A070504	Premature Discharge of Battery	03/04/2019	
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	03/04/2019	
					CT449	Generators and associated devices	F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A070504	Premature Discharge of Battery	03/04/2019	
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	03/04/2019	
					CT981	Single use	F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A070504	Premature Discharge of Battery	03/04/2019	
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	03/04/2019	
					CT983	Surgical invasive	F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A070504	Premature Discharge of Battery	03/04/2019	
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	03/04/2019	
8826	2019	All other reportable incidents	47265	Dual-chamber implantable pacemaker, rate-responsive	CT141	Electrodes/Leads and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
					CT145	Cardiovascular devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
					CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
					CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
					CT233	Surgical	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
					CT2406	Implantable	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
					CT2407	Active implantable	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
					CT2727	Active implant devices/ancillaries	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
					CT2750	Implantable cardiac electrical stimulators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
					CT2985	Active	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
					CT3001	Implantable metal	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
					CT3002	Implantable synthetic polymer	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
					CT3006	Implantable inorganic material	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
					CT314	Electrically powered	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
					CT321	Long-term surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
					CT449	Generators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
					CT981	Single use	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
					CT983	Surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
8827	2019	All other reportable incidents	36241	Pacemaker/defibrillator lead	CT1684	Cardiac pacemakers and associated devices	F1905	Device Revision or Replacement	A0510	Retraction Problem	05/04/2019	
8828	2019	All other reportable incidents	47267	Single-chamber implantable pacemaker, rate-	CT141	Electrodes/Leads and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	12/04/2019	

				responsive	CT145	Cardiovascular devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	12/04/2019		
					CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	12/04/2019		
					CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	12/04/2019		
					CT233	Surgical	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	12/04/2019		
					CT2406	Implantable	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	12/04/2019		
					CT2407	Active implantable	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	12/04/2019		
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	12/04/2019		
					CT2727	Active implant devices/ancillaries	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	12/04/2019		
					CT2750	Implantable cardiac electrical stimulators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	12/04/2019		
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	12/04/2019		
					CT2985	Active	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	12/04/2019		
					CT3001	Implantable metal	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	12/04/2019		
					CT3002	Implantable synthetic polymer	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	12/04/2019		
					CT3006	Implantable inorganic material	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	12/04/2019		
					CT314	Electrically powered	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	12/04/2019		
					CT321	Long-term surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	12/04/2019		
					CT449	Generators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	12/04/2019		
					CT981	Single use	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	12/04/2019		
					CT983	Surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	12/04/2019		
8829	2019	All other reportable incidents	36241	Pacemaker/defibrillator lead	CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A0510	Retraction Problem	27/03/2019		
8830	2019	All other reportable incidents	36241	Pacemaker/defibrillator lead	CT1684	Cardiac pacemakers and associated devices	F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A070909	Oversensing			
8831	2019	All other reportable incidents	35852	Single-chamber implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A070909	Oversensing			
8832	2019	All other reportable incidents	37265	Dual-chamber implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A072201	High Impedance			
8833	2019	All other reportable incidents	37265	Dual-chamber implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A070909	Oversensing			
8834	2019	All other reportable incidents	36241	Pacemaker/defibrillator lead	CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A072202	Low Impedance	04/04/2019		
8835	2019	All other reportable incidents	35852	Single-chamber implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A070910	Undersensing			
8836	2019	All other reportable incidents	35852	Single-chamber implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A071302	Inappropriate/Inadequate Shock/Stimulation			
8837	2019	All other reportable incidents	36241	Pacemaker/defibrillator lead	CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A072201	High Impedance			
8838	2019	All other reportable incidents	36241	Pacemaker/defibrillator lead	CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A072201	High Impedance			
8839	2019	All other reportable incidents	35852	Single-chamber implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A072201	High Impedance			
8840	2020	All other reportable incidents	37265	Dual-chamber implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A160601	Premature Elective Replacement Indicator			
							F1903	Device Explantation	A160601	Premature Elective Replacement Indicator			
							F1905	Device Revision or Replacement	A160601	Premature Elective Replacement Indicator			
8841	2019	All other reportable incidents	36241	Pacemaker/defibrillator lead	CT1684	Cardiac pacemakers and associated devices	F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A	Annex A			
8842	2019	All other reportable incidents	35852	Single-chamber implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A071302	Inappropriate/Inadequate Shock/Stimulation			
8843	2020	All other reportable incidents	35852	Single-chamber implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E2104	Electric Shock	A071302	Inappropriate/Inadequate Shock/Stimulation	01/03/2021	01/03/2021	
							F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A071302	Inappropriate/Inadequate Shock/Stimulation	01/03/2021	01/03/2021	
							F1905	Device Revision or Replacement	A071302	Inappropriate/Inadequate Shock/Stimulation	01/03/2021	01/03/2021	
8844	2019	All other reportable incidents	35223	Endocardial/interventricular septal pacing lead	CT1684	Cardiac pacemakers and associated devices	F1905	Device Revision or Replacement	A0709	Device Sensing Problem	01/03/2019		
8845	2020	Death	35223	Endocardial/interventricular septal pacing lead	CT1684	Cardiac pacemakers and associated devices	E0611	Heart Failure/Congestive Heart Failure	A24	No Apparent Device Problem	19/01/2020	19/01/2020	
							F02	Death	A24	No Apparent Device Problem	19/01/2020	19/01/2020	
8846	2019	All other reportable incidents	37265	Dual-chamber implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A	Annex A	09/04/2019		
							F1905	Device Revision or Replacement	A	Annex A	09/04/2019		
8847	2020	All other reportable incidents	17882	Physiologic-monitoring defibrillation system	CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A0719	Unexpected Shutdown			
							F26	No Health Consequences or Impact	A0719	Unexpected Shutdown			
8848	2020	All other reportable incidents	37265	Dual-chamber implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	07/02/2020	07/02/2020	
8849	2019	All other reportable incidents	47270	Cardiac resynchronization therapy implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery			
8850	2019	All other reportable incidents	35852	Single-chamber implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A070909	Oversensing			
8851	2019	All other reportable incidents	17882	Physiologic-monitoring defibrillation system	CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem			
8852	2020	All other reportable incidents	37265	Dual-chamber implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	02/03/2020	02/03/2020	
							F1901	Additional Surgery	A070504	Premature Discharge of Battery	02/03/2020	02/03/2020	
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	02/03/2020	02/03/2020	
8853	2020	All other reportable incidents	47270	Cardiac resynchronization therapy implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	25/02/2020	25/02/2020	
							F1901	Additional Surgery	A070504	Premature Discharge of Battery	25/02/2020	25/02/2020	
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	25/02/2020	25/02/2020	
8854	2020	All other reportable incidents	36241	Pacemaker/defibrillator lead	CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A072201	High Impedance			
							F26	No Health Consequences or Impact	A072201	High Impedance			
8855	2020	All other reportable incidents	47265	Dual-chamber implantable pacemaker, rate-responsive	CT141	Electrodes/Leads and associated devices	F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	24/02/2020	24/02/2020	
					CT145	Cardiovascular devices	F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	24/02/2020	24/02/2020	
					CT1684	Cardiac pacemakers and associated devices	F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	24/02/2020	24/02/2020	
					CT1685	Cardiac pacemakers	F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	24/02/2020	24/02/2020	
					CT233	Surgical	F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	24/02/2020	24/02/2020	
					CT2406	Implantable	F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	24/02/2020	24/02/2020	
					CT2407	Active implantable	F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	24/02/2020	24/02/2020	
					CT257	Implantable cardiac pacemakers	F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	24/02/2020	24/02/2020	
					CT2727	Active implant devices/ancillaries	F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	24/02/2020	24/02/2020	
					CT2750	Implantable cardiac electrical stimulators and associated devices	F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	24/02/2020	24/02/2020	
					CT2866	Implantable cardiac electrical stimulators	F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	24/02/2020	24/02/2020	
					CT2985	Active	F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	24/02/2020	24/02/2020	
					CT3001	Implantable metal	F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	24/02/2020	24/02/2020	
					CT3002	Implantable synthetic polymer	F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	24/02/2020	24/02/2020	
					CT3006	Implantable inorganic material	F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	24/02/2020	24/02/2020	
					CT314	Electrically powered	F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	24/02/2020	24/02/2020	
					CT321	Long-term surgical invasive	F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	24/02/2020	24/02/2020	
					CT449	Generators and associated devices	F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	24/02/2020	24/02/2020	
					CT981	Single use	F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	24/02/2020	24/02/2020	
					CT983	Surgical invasive	F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	24/02/2020	24/02/2020	
8856	2020	All other reportable incidents	35852	Single-chamber implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E2104	Electric Shock	A071302	Inappropriate/Inadequate Shock/Stimulation			
8857	2020	All other reportable incidents	35852	Single-chamber implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	24/02/2021	24/02/2021	
							F19	Surgical Intervention	A070504	Premature Discharge of Battery	24/02/2021	24/02/2021	
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	24/02/2021	24/02/2021	
8858	2019	All other reportable incidents	47263	Cardiac resynchronization therapy implantable pacemaker	CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070101	Failure to Capture			
					CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A070101	Failure to Capture			
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A070101	Failure to Capture			
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A070101	Failure to Capture			

8859	2019	All other reportable incidents	47270	Cardiac resynchronization therapy implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery			
8860	2020	All other reportable incidents	35223	Endocardial/interventricular septal pacing lead	CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A090801	Signal Artifact/Noise	09/01/2019	09/01/2019	
							F1901	Additional Surgery	A090801	Signal Artifact/Noise	09/01/2019	09/01/2019	
							F1905	Device Revision or Replacement	A090801	Signal Artifact/Noise	09/01/2019	09/01/2019	
8861	2019	All other reportable incidents	35223	Endocardial/interventricular septal pacing lead	CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070909	Oversensing			
8862	2020	All other reportable incidents	35223	Endocardial/interventricular septal pacing lead	CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A090801	Signal Artifact/Noise	09/01/2019	09/01/2019	
							F1901	Additional Surgery	A090801	Signal Artifact/Noise	09/01/2019	09/01/2019	
							F1905	Device Revision or Replacement	A090801	Signal Artifact/Noise	09/01/2019	09/01/2019	
8863	2019	All other reportable incidents	47265	Dual-chamber implantable pacemaker, rate-responsive	CT141	Electrodes/Leads and associated devices	F1905	Device Revision or Replacement	A0721	Circuit Failure	10/05/2019		
					CT145	Cardiovascular devices	F1905	Device Revision or Replacement	A0721	Circuit Failure	10/05/2019		
					CT1684	Cardiac pacemakers and associated devices	F1905	Device Revision or Replacement	A0721	Circuit Failure	10/05/2019		
					CT1685	Cardiac pacemakers	F1905	Device Revision or Replacement	A0721	Circuit Failure	10/05/2019		
					CT233	Surgical	F1905	Device Revision or Replacement	A0721	Circuit Failure	10/05/2019		
					CT2406	Implantable	F1905	Device Revision or Replacement	A0721	Circuit Failure	10/05/2019		
					CT2407	Active implantable	F1905	Device Revision or Replacement	A0721	Circuit Failure	10/05/2019		
					CT257	Implantable cardiac pacemakers	F1905	Device Revision or Replacement	A0721	Circuit Failure	10/05/2019		
					CT2727	Active implant devices/ancillaries	F1905	Device Revision or Replacement	A0721	Circuit Failure	10/05/2019		
					CT2750	Implantable cardiac electrical stimulators and associated devices	F1905	Device Revision or Replacement	A0721	Circuit Failure	10/05/2019		
					CT2866	Implantable cardiac electrical stimulators	F1905	Device Revision or Replacement	A0721	Circuit Failure	10/05/2019		
					CT2985	Active	F1905	Device Revision or Replacement	A0721	Circuit Failure	10/05/2019		
					CT3001	Implantable metal	F1905	Device Revision or Replacement	A0721	Circuit Failure	10/05/2019		
					CT3002	Implantable synthetic polymer	F1905	Device Revision or Replacement	A0721	Circuit Failure	10/05/2019		
					CT3006	Implantable inorganic material	F1905	Device Revision or Replacement	A0721	Circuit Failure	10/05/2019		
					CT314	Electrically powered	F1905	Device Revision or Replacement	A0721	Circuit Failure	10/05/2019		
					CT321	Long-term surgical invasive	F1905	Device Revision or Replacement	A0721	Circuit Failure	10/05/2019		
					CT449	Generators and associated devices	F1905	Device Revision or Replacement	A0721	Circuit Failure	10/05/2019		
					CT981	Single use	F1905	Device Revision or Replacement	A0721	Circuit Failure	10/05/2019		
					CT983	Surgical invasive	F1905	Device Revision or Replacement	A0721	Circuit Failure	10/05/2019		
8864	2019	All other reportable incidents	47265	Dual-chamber implantable pacemaker, rate-responsive	CT141	Electrodes/Leads and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A072201	High Impedance			
					CT145	Cardiovascular devices	E2403	No Clinical Signs, Symptoms or Conditions	A072201	High Impedance			
					CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A072201	High Impedance			
					CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A072201	High Impedance			
					CT233	Surgical	E2403	No Clinical Signs, Symptoms or Conditions	A072201	High Impedance			
					CT2406	Implantable	E2403	No Clinical Signs, Symptoms or Conditions	A072201	High Impedance			
					CT2407	Active implantable	E2403	No Clinical Signs, Symptoms or Conditions	A072201	High Impedance			
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A072201	High Impedance			
					CT2727	Active implant devices/ancillaries	E2403	No Clinical Signs, Symptoms or Conditions	A072201	High Impedance			
					CT2750	Implantable cardiac electrical stimulators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A072201	High Impedance			
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A072201	High Impedance			
					CT2985	Active	E2403	No Clinical Signs, Symptoms or Conditions	A072201	High Impedance			
					CT3001	Implantable metal	E2403	No Clinical Signs, Symptoms or Conditions	A072201	High Impedance			
					CT3002	Implantable synthetic polymer	E2403	No Clinical Signs, Symptoms or Conditions	A072201	High Impedance			
					CT3006	Implantable inorganic material	E2403	No Clinical Signs, Symptoms or Conditions	A072201	High Impedance			
					CT314	Electrically powered	E2403	No Clinical Signs, Symptoms or Conditions	A072201	High Impedance			
					CT321	Long-term surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A072201	High Impedance			
					CT449	Generators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A072201	High Impedance			
					CT981	Single use	E2403	No Clinical Signs, Symptoms or Conditions	A072201	High Impedance			
					CT983	Surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A072201	High Impedance			
8865	2020	All other reportable incidents	37265	Dual-chamber implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	01/03/2020	01/03/2020	
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	01/03/2020	01/03/2020	
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	01/03/2020	01/03/2020	
8866	2020	All other reportable incidents	35223	Endocardial/interventricular septal pacing lead	CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070909	Oversensing	12/02/2020	12/02/2020	
									A071204	No Pacing	12/02/2020	12/02/2020	
									A150204	Failure to Advance	12/02/2020	12/02/2020	
							F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A070909	Oversensing	12/02/2020	12/02/2020	
									A071204	No Pacing	12/02/2020	12/02/2020	
									A150204	Failure to Advance	12/02/2020	12/02/2020	
							F1905	Device Revision or Replacement	A070909	Oversensing	12/02/2020	12/02/2020	
									A071204	No Pacing	12/02/2020	12/02/2020	
									A150204	Failure to Advance	12/02/2020	12/02/2020	
							F1908	Prolonged Surgery	A070909	Oversensing	12/02/2020	12/02/2020	
									A071204	No Pacing	12/02/2020	12/02/2020	
									A150204	Failure to Advance	12/02/2020	12/02/2020	
8867	2020	All other reportable incidents	36241	Pacemaker/defibrillator lead	CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A051201	Device Dislodged or Dislocated	07/03/2019	07/03/2019	
									A0709	Device Sensing Problem	07/03/2019	07/03/2019	
									A0712	Pacing Problem	07/03/2019	07/03/2019	
8868	2020	All other reportable incidents	35852	Single-chamber implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E2104	Electric Shock	A0712	Pacing Problem			
									A071302	Inappropriate/Inadequate Shock/Stimulation			
							F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A0712	Pacing Problem			
									A071302	Inappropriate/Inadequate Shock/Stimulation			
							F26	No Health Consequences or Impact	A0712	Pacing Problem			
									A071302	Inappropriate/Inadequate Shock/Stimulation			
8869	2020	All other reportable incidents	35852	Single-chamber implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A070909	Oversensing			
									A071302	Inappropriate/Inadequate Shock/Stimulation			
8870	2020	All other reportable incidents	35852	Single-chamber implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E2104	Electric Shock	A070909	Oversensing			
									A071302	Inappropriate/Inadequate Shock/Stimulation			
							F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A070909	Oversensing			
									A071302	Inappropriate/Inadequate Shock/Stimulation			
							F23	Unexpected Medical Intervention	A070909	Oversensing			
									A071302	Inappropriate/Inadequate Shock/Stimulation			
8871	2019	All other reportable incidents	36241	Pacemaker/defibrillator lead	CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070101	Failure to Capture			
8872	2019	All other reportable incidents	47265	Dual-chamber implantable pacemaker, rate-responsive	CT141	Electrodes/Leads and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070101	Failure to Capture			
					CT145	Cardiovascular devices	E2403	No Clinical Signs, Symptoms or Conditions	A070101	Failure to Capture			
					CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070101	Failure to Capture			
					CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A070101	Failure to Capture			

					CT233	Surgical	E2403	No Clinical Signs, Symptoms or Conditions	A070101	Failure to Capture			
					CT2406	Implantable	E2403	No Clinical Signs, Symptoms or Conditions	A070101	Failure to Capture			
					CT2407	Active implantable	E2403	No Clinical Signs, Symptoms or Conditions	A070101	Failure to Capture			
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A070101	Failure to Capture			
					CT2727	Active implant devices/ancillaries	E2403	No Clinical Signs, Symptoms or Conditions	A070101	Failure to Capture			
					CT2750	Implantable cardiac electrical stimulators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070101	Failure to Capture			
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A070101	Failure to Capture			
					CT2985	Active	E2403	No Clinical Signs, Symptoms or Conditions	A070101	Failure to Capture			
					CT3001	Implantable metal	E2403	No Clinical Signs, Symptoms or Conditions	A070101	Failure to Capture			
					CT3002	Implantable synthetic polymer	E2403	No Clinical Signs, Symptoms or Conditions	A070101	Failure to Capture			
					CT3006	Implantable inorganic material	E2403	No Clinical Signs, Symptoms or Conditions	A070101	Failure to Capture			
					CT314	Electrically powered	E2403	No Clinical Signs, Symptoms or Conditions	A070101	Failure to Capture			
					CT321	Long-term surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A070101	Failure to Capture			
					CT449	Generators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070101	Failure to Capture			
					CT981	Single use	E2403	No Clinical Signs, Symptoms or Conditions	A070101	Failure to Capture			
					CT983	Surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A070101	Failure to Capture			
8873	2019	All other reportable incidents	36241	Pacemaker/defibrillator lead	CT1684	Cardiac pacemakers and associated devices	F1905	Device Revision or Replacement	A0510	Retraction Problem	09/04/2019		
8874	2019	All other reportable incidents	36241	Pacemaker/defibrillator lead	CT1684	Cardiac pacemakers and associated devices	F05	Delay to Treatment/ Therapy	A0510	Retraction Problem	09/04/2019		
8875	2020	All other reportable incidents	35223	Endocardial/interventricular septal pacing lead	CT1684	Cardiac pacemakers and associated devices	E0604	Cardiac Perforation	A070101	Failure to Capture	29/01/2020	29/01/2020	
									A24	No Apparent Device Problem	29/01/2020	29/01/2020	
							F12	Serious Injury/ Illness/ Impairment	A070101	Failure to Capture	29/01/2020	29/01/2020	
									A24	No Apparent Device Problem	29/01/2020	29/01/2020	
							F1905	Device Revision or Replacement	A070101	Failure to Capture	29/01/2020	29/01/2020	
									A24	No Apparent Device Problem	29/01/2020	29/01/2020	
8876	2020	All other reportable incidents	47267	Single-chamber implantable pacemaker, rate-responsive	CT141	Electrodes/Leads and associated devices	F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	17/02/2020	17/02/2020	
					CT145	Cardiovascular devices	F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	17/02/2020	17/02/2020	
					CT1684	Cardiac pacemakers and associated devices	F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	17/02/2020	17/02/2020	
					CT1685	Cardiac pacemakers	F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	17/02/2020	17/02/2020	
					CT233	Surgical	F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	17/02/2020	17/02/2020	
					CT2406	Implantable	F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	17/02/2020	17/02/2020	
					CT2407	Active implantable	F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	17/02/2020	17/02/2020	
					CT257	Implantable cardiac pacemakers	F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	17/02/2020	17/02/2020	
					CT2727	Active implant devices/ancillaries	F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	17/02/2020	17/02/2020	
					CT2750	Implantable cardiac electrical stimulators and associated devices	F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	17/02/2020	17/02/2020	
					CT2866	Implantable cardiac electrical stimulators	F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	17/02/2020	17/02/2020	
					CT2985	Active	F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	17/02/2020	17/02/2020	
					CT3001	Implantable metal	F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	17/02/2020	17/02/2020	
					CT3002	Implantable synthetic polymer	F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	17/02/2020	17/02/2020	
					CT3006	Implantable inorganic material	F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	17/02/2020	17/02/2020	
					CT314	Electrically powered	F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	17/02/2020	17/02/2020	
					CT321	Long-term surgical invasive	F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	17/02/2020	17/02/2020	
					CT449	Generators and associated devices	F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	17/02/2020	17/02/2020	
					CT981	Single use	F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	17/02/2020	17/02/2020	
					CT983	Surgical invasive	F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	17/02/2020	17/02/2020	
8877	2020	All other reportable incidents	47265	Dual-chamber implantable pacemaker, rate-responsive	CT141	Electrodes/Leads and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	03/04/2020	03/04/2020	
							F19	Surgical Intervention	A070504	Premature Discharge of Battery	03/04/2020	03/04/2020	
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	03/04/2020	03/04/2020	
					CT145	Cardiovascular devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	03/04/2020	03/04/2020	
							F19	Surgical Intervention	A070504	Premature Discharge of Battery	03/04/2020	03/04/2020	
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	03/04/2020	03/04/2020	
					CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	03/04/2020	03/04/2020	
							F19	Surgical Intervention	A070504	Premature Discharge of Battery	03/04/2020	03/04/2020	
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	03/04/2020	03/04/2020	
					CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	03/04/2020	03/04/2020	
							F19	Surgical Intervention	A070504	Premature Discharge of Battery	03/04/2020	03/04/2020	
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	03/04/2020	03/04/2020	
					CT233	Surgical	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	03/04/2020	03/04/2020	
							F19	Surgical Intervention	A070504	Premature Discharge of Battery	03/04/2020	03/04/2020	
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	03/04/2020	03/04/2020	
					CT2406	Implantable	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	03/04/2020	03/04/2020	
							F19	Surgical Intervention	A070504	Premature Discharge of Battery	03/04/2020	03/04/2020	
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	03/04/2020	03/04/2020	
					CT2407	Active implantable	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	03/04/2020	03/04/2020	
							F19	Surgical Intervention	A070504	Premature Discharge of Battery	03/04/2020	03/04/2020	
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	03/04/2020	03/04/2020	
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	03/04/2020	03/04/2020	
							F19	Surgical Intervention	A070504	Premature Discharge of Battery	03/04/2020	03/04/2020	
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	03/04/2020	03/04/2020	
					CT2727	Active implant devices/ancillaries	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	03/04/2020	03/04/2020	
							F19	Surgical Intervention	A070504	Premature Discharge of Battery	03/04/2020	03/04/2020	
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	03/04/2020	03/04/2020	
					CT2750	Implantable cardiac electrical stimulators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	03/04/2020	03/04/2020	
							F19	Surgical Intervention	A070504	Premature Discharge of Battery	03/04/2020	03/04/2020	
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	03/04/2020	03/04/2020	
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	03/04/2020	03/04/2020	
							F19	Surgical Intervention	A070504	Premature Discharge of Battery	03/04/2020	03/04/2020	
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	03/04/2020	03/04/2020	
					CT2985	Active	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	03/04/2020	03/04/2020	
							F19	Surgical Intervention	A070504	Premature Discharge of Battery	03/04/2020	03/04/2020	
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	03/04/2020	03/04/2020	
					CT3001	Implantable metal	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	03/04/2020	03/04/2020	
							F19	Surgical Intervention	A070504	Premature Discharge of Battery	03/04/2020	03/04/2020	
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	03/04/2020	03/04/2020	
					CT3002	Implantable synthetic polymer	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	03/04/2020	03/04/2020	
							F19	Surgical Intervention	A070504	Premature Discharge of Battery	03/04/2020	03/04/2020	
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	03/04/2020	03/04/2020	

					CT3006	Implantable inorganic material	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	03/04/2020	03/04/2020
							F19	Surgical Intervention	A070504	Premature Discharge of Battery	03/04/2020	03/04/2020
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	03/04/2020	03/04/2020
					CT314	Electrically powered	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	03/04/2020	03/04/2020
							F19	Surgical Intervention	A070504	Premature Discharge of Battery	03/04/2020	03/04/2020
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	03/04/2020	03/04/2020
					CT321	Long-term surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	03/04/2020	03/04/2020
							F19	Surgical Intervention	A070504	Premature Discharge of Battery	03/04/2020	03/04/2020
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	03/04/2020	03/04/2020
					CT449	Generators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	03/04/2020	03/04/2020
							F19	Surgical Intervention	A070504	Premature Discharge of Battery	03/04/2020	03/04/2020
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	03/04/2020	03/04/2020
					CT981	Single use	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	03/04/2020	03/04/2020
							F19	Surgical Intervention	A070504	Premature Discharge of Battery	03/04/2020	03/04/2020
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	03/04/2020	03/04/2020
					CT983	Surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	03/04/2020	03/04/2020
							F19	Surgical Intervention	A070504	Premature Discharge of Battery	03/04/2020	03/04/2020
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	03/04/2020	03/04/2020
8878	2020	All other reportable incidents	35852	Single-chamber implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E060109	Tachycardia	A070910	Undersensing		
8879	2020	All other reportable incidents	36241	Pacemaker/defibrillator lead	CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A051201	Device Dislodged or Dislocated		
8880	2020	All other reportable incidents	47265	Dual-chamber implantable pacemaker, rate-responsive	CT141	Electrodes/Leads and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070101	Failure to Capture		
							F26	No Health Consequences or Impact	A070101	Failure to Capture		
					CT145	Cardiovascular devices	E2403	No Clinical Signs, Symptoms or Conditions	A070101	Failure to Capture		
							F26	No Health Consequences or Impact	A070101	Failure to Capture		
					CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070101	Failure to Capture		
							F26	No Health Consequences or Impact	A070101	Failure to Capture		
					CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A070101	Failure to Capture		
							F26	No Health Consequences or Impact	A070101	Failure to Capture		
					CT233	Surgical	E2403	No Clinical Signs, Symptoms or Conditions	A070101	Failure to Capture		
							F26	No Health Consequences or Impact	A070101	Failure to Capture		
					CT2406	Implantable	E2403	No Clinical Signs, Symptoms or Conditions	A070101	Failure to Capture		
							F26	No Health Consequences or Impact	A070101	Failure to Capture		
					CT2407	Active implantable	E2403	No Clinical Signs, Symptoms or Conditions	A070101	Failure to Capture		
							F26	No Health Consequences or Impact	A070101	Failure to Capture		
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A070101	Failure to Capture		
							F26	No Health Consequences or Impact	A070101	Failure to Capture		
					CT2727	Active implant devices/ancillaries	E2403	No Clinical Signs, Symptoms or Conditions	A070101	Failure to Capture		
							F26	No Health Consequences or Impact	A070101	Failure to Capture		
					CT2750	Implantable cardiac electrical stimulators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070101	Failure to Capture		
							F26	No Health Consequences or Impact	A070101	Failure to Capture		
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A070101	Failure to Capture		
							F26	No Health Consequences or Impact	A070101	Failure to Capture		
					CT2985	Active	E2403	No Clinical Signs, Symptoms or Conditions	A070101	Failure to Capture		
							F26	No Health Consequences or Impact	A070101	Failure to Capture		
					CT3001	Implantable metal	E2403	No Clinical Signs, Symptoms or Conditions	A070101	Failure to Capture		
							F26	No Health Consequences or Impact	A070101	Failure to Capture		
					CT3002	Implantable synthetic polymer	E2403	No Clinical Signs, Symptoms or Conditions	A070101	Failure to Capture		
							F26	No Health Consequences or Impact	A070101	Failure to Capture		
					CT3006	Implantable inorganic material	E2403	No Clinical Signs, Symptoms or Conditions	A070101	Failure to Capture		
							F26	No Health Consequences or Impact	A070101	Failure to Capture		
					CT314	Electrically powered	E2403	No Clinical Signs, Symptoms or Conditions	A070101	Failure to Capture		
							F26	No Health Consequences or Impact	A070101	Failure to Capture		
					CT321	Long-term surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A070101	Failure to Capture		
							F26	No Health Consequences or Impact	A070101	Failure to Capture		
					CT449	Generators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070101	Failure to Capture		
							F26	No Health Consequences or Impact	A070101	Failure to Capture		
					CT981	Single use	E2403	No Clinical Signs, Symptoms or Conditions	A070101	Failure to Capture		
							F26	No Health Consequences or Impact	A070101	Failure to Capture		
					CT983	Surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A070101	Failure to Capture		
							F26	No Health Consequences or Impact	A070101	Failure to Capture		
8881	2020	All other reportable incidents	37265	Dual-chamber implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	13/03/2020	13/03/2020
							F1901	Additional Surgery	A070504	Premature Discharge of Battery	13/03/2020	13/03/2020
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	13/03/2020	13/03/2020
8882	2020	All other reportable incidents	47265	Dual-chamber implantable pacemaker, rate-responsive	CT141	Electrodes/Leads and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem	26/03/2020	26/03/2020
									A070504	Premature Discharge of Battery	26/03/2020	26/03/2020
							F08	Hospitalization or Prolonged Hospitalization	A0705	Battery Problem	26/03/2020	26/03/2020
									A070504	Premature Discharge of Battery	26/03/2020	26/03/2020
							F12	Serious Injury/ Illness/ Impairment	A0705	Battery Problem	26/03/2020	26/03/2020
									A070504	Premature Discharge of Battery	26/03/2020	26/03/2020
							F19	Surgical Intervention	A0705	Battery Problem	26/03/2020	26/03/2020
									A070504	Premature Discharge of Battery	26/03/2020	26/03/2020
							F1905	Device Revision or Replacement	A0705	Battery Problem	26/03/2020	26/03/2020
									A070504	Premature Discharge of Battery	26/03/2020	26/03/2020
					CT145	Cardiovascular devices	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem	26/03/2020	26/03/2020
									A070504	Premature Discharge of Battery	26/03/2020	26/03/2020
							F08	Hospitalization or Prolonged Hospitalization	A0705	Battery Problem	26/03/2020	26/03/2020
									A070504	Premature Discharge of Battery	26/03/2020	26/03/2020
							F12	Serious Injury/ Illness/ Impairment	A0705	Battery Problem	26/03/2020	26/03/2020
									A070504	Premature Discharge of Battery	26/03/2020	26/03/2020
							F19	Surgical Intervention	A0705	Battery Problem	26/03/2020	26/03/2020
									A070504	Premature Discharge of Battery	26/03/2020	26/03/2020
							F1905	Device Revision or Replacement	A0705	Battery Problem	26/03/2020	26/03/2020
									A070504	Premature Discharge of Battery	26/03/2020	26/03/2020
					CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem	26/03/2020	26/03/2020
									A070504	Premature Discharge of Battery	26/03/2020	26/03/2020
							F08	Hospitalization or Prolonged Hospitalization	A0705	Battery Problem	26/03/2020	26/03/2020
									A070504	Premature Discharge of Battery	26/03/2020	26/03/2020
							F12	Serious Injury/ Illness/ Impairment	A0705	Battery Problem	26/03/2020	26/03/2020

									A070504	Premature Discharge of Battery	26/03/2020	26/03/2020
							F19	Surgical Intervention	A0705	Battery Problem	26/03/2020	26/03/2020
									A070504	Premature Discharge of Battery	26/03/2020	26/03/2020
							F1905	Device Revision or Replacement	A0705	Battery Problem	26/03/2020	26/03/2020
									A070504	Premature Discharge of Battery	26/03/2020	26/03/2020
					CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem	26/03/2020	26/03/2020
									A070504	Premature Discharge of Battery	26/03/2020	26/03/2020
							F08	Hospitalization or Prolonged Hospitalization	A0705	Battery Problem	26/03/2020	26/03/2020
									A070504	Premature Discharge of Battery	26/03/2020	26/03/2020
							F12	Serious Injury/ Illness/ Impairment	A0705	Battery Problem	26/03/2020	26/03/2020
									A070504	Premature Discharge of Battery	26/03/2020	26/03/2020
							F19	Surgical Intervention	A0705	Battery Problem	26/03/2020	26/03/2020
									A070504	Premature Discharge of Battery	26/03/2020	26/03/2020
							F1905	Device Revision or Replacement	A0705	Battery Problem	26/03/2020	26/03/2020
									A070504	Premature Discharge of Battery	26/03/2020	26/03/2020
					CT233	Surgical	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem	26/03/2020	26/03/2020
									A070504	Premature Discharge of Battery	26/03/2020	26/03/2020
							F08	Hospitalization or Prolonged Hospitalization	A0705	Battery Problem	26/03/2020	26/03/2020
									A070504	Premature Discharge of Battery	26/03/2020	26/03/2020
							F12	Serious Injury/ Illness/ Impairment	A0705	Battery Problem	26/03/2020	26/03/2020
									A070504	Premature Discharge of Battery	26/03/2020	26/03/2020
							F19	Surgical Intervention	A0705	Battery Problem	26/03/2020	26/03/2020
									A070504	Premature Discharge of Battery	26/03/2020	26/03/2020
							F1905	Device Revision or Replacement	A0705	Battery Problem	26/03/2020	26/03/2020
									A070504	Premature Discharge of Battery	26/03/2020	26/03/2020
					CT2406	Implantable	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem	26/03/2020	26/03/2020
									A070504	Premature Discharge of Battery	26/03/2020	26/03/2020
							F08	Hospitalization or Prolonged Hospitalization	A0705	Battery Problem	26/03/2020	26/03/2020
									A070504	Premature Discharge of Battery	26/03/2020	26/03/2020
							F12	Serious Injury/ Illness/ Impairment	A0705	Battery Problem	26/03/2020	26/03/2020
									A070504	Premature Discharge of Battery	26/03/2020	26/03/2020
							F19	Surgical Intervention	A0705	Battery Problem	26/03/2020	26/03/2020
									A070504	Premature Discharge of Battery	26/03/2020	26/03/2020
							F1905	Device Revision or Replacement	A0705	Battery Problem	26/03/2020	26/03/2020
									A070504	Premature Discharge of Battery	26/03/2020	26/03/2020
					CT2407	Active implantable	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem	26/03/2020	26/03/2020
									A070504	Premature Discharge of Battery	26/03/2020	26/03/2020
							F08	Hospitalization or Prolonged Hospitalization	A0705	Battery Problem	26/03/2020	26/03/2020
									A070504	Premature Discharge of Battery	26/03/2020	26/03/2020
							F12	Serious Injury/ Illness/ Impairment	A0705	Battery Problem	26/03/2020	26/03/2020
									A070504	Premature Discharge of Battery	26/03/2020	26/03/2020
							F19	Surgical Intervention	A0705	Battery Problem	26/03/2020	26/03/2020
									A070504	Premature Discharge of Battery	26/03/2020	26/03/2020
							F1905	Device Revision or Replacement	A0705	Battery Problem	26/03/2020	26/03/2020
									A070504	Premature Discharge of Battery	26/03/2020	26/03/2020
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem	26/03/2020	26/03/2020
									A070504	Premature Discharge of Battery	26/03/2020	26/03/2020
							F08	Hospitalization or Prolonged Hospitalization	A0705	Battery Problem	26/03/2020	26/03/2020
									A070504	Premature Discharge of Battery	26/03/2020	26/03/2020
							F12	Serious Injury/ Illness/ Impairment	A0705	Battery Problem	26/03/2020	26/03/2020
									A070504	Premature Discharge of Battery	26/03/2020	26/03/2020
							F19	Surgical Intervention	A0705	Battery Problem	26/03/2020	26/03/2020
									A070504	Premature Discharge of Battery	26/03/2020	26/03/2020
							F1905	Device Revision or Replacement	A0705	Battery Problem	26/03/2020	26/03/2020
									A070504	Premature Discharge of Battery	26/03/2020	26/03/2020
					CT2727	Active implant devices/ancillaries	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem	26/03/2020	26/03/2020
									A070504	Premature Discharge of Battery	26/03/2020	26/03/2020
							F08	Hospitalization or Prolonged Hospitalization	A0705	Battery Problem	26/03/2020	26/03/2020
									A070504	Premature Discharge of Battery	26/03/2020	26/03/2020
							F12	Serious Injury/ Illness/ Impairment	A0705	Battery Problem	26/03/2020	26/03/2020
									A070504	Premature Discharge of Battery	26/03/2020	26/03/2020
							F19	Surgical Intervention	A0705	Battery Problem	26/03/2020	26/03/2020
									A070504	Premature Discharge of Battery	26/03/2020	26/03/2020
							F1905	Device Revision or Replacement	A0705	Battery Problem	26/03/2020	26/03/2020
									A070504	Premature Discharge of Battery	26/03/2020	26/03/2020
					CT2750	Implantable cardiac electrical stimulators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem	26/03/2020	26/03/2020
									A070504	Premature Discharge of Battery	26/03/2020	26/03/2020
							F08	Hospitalization or Prolonged Hospitalization	A0705	Battery Problem	26/03/2020	26/03/2020
									A070504	Premature Discharge of Battery	26/03/2020	26/03/2020
							F12	Serious Injury/ Illness/ Impairment	A0705	Battery Problem	26/03/2020	26/03/2020
									A070504	Premature Discharge of Battery	26/03/2020	26/03/2020
							F19	Surgical Intervention	A0705	Battery Problem	26/03/2020	26/03/2020
									A070504	Premature Discharge of Battery	26/03/2020	26/03/2020
							F1905	Device Revision or Replacement	A0705	Battery Problem	26/03/2020	26/03/2020
									A070504	Premature Discharge of Battery	26/03/2020	26/03/2020
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem	26/03/2020	26/03/2020
									A070504	Premature Discharge of Battery	26/03/2020	26/03/2020
							F08	Hospitalization or Prolonged Hospitalization	A0705	Battery Problem	26/03/2020	26/03/2020
									A070504	Premature Discharge of Battery	26/03/2020	26/03/2020
							F12	Serious Injury/ Illness/ Impairment	A0705	Battery Problem	26/03/2020	26/03/2020
									A070504	Premature Discharge of Battery	26/03/2020	26/03/2020
							F19	Surgical Intervention	A0705	Battery Problem	26/03/2020	26/03/2020
									A070504	Premature Discharge of Battery	26/03/2020	26/03/2020
							F1905	Device Revision or Replacement	A0705	Battery Problem	26/03/2020	26/03/2020
									A070504	Premature Discharge of Battery	26/03/2020	26/03/2020
					CT2985	Active	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem	26/03/2020	26/03/2020
									A070504	Premature Discharge of Battery	26/03/2020	26/03/2020
							F08	Hospitalization or Prolonged Hospitalization	A0705	Battery Problem	26/03/2020	26/03/2020

									A070504	Premature Discharge of Battery	26/03/2020	26/03/2020
							F12	Serious Injury/ Illness/ Impairment	A0705	Battery Problem	26/03/2020	26/03/2020
									A070504	Premature Discharge of Battery	26/03/2020	26/03/2020
							F19	Surgical Intervention	A0705	Battery Problem	26/03/2020	26/03/2020
									A070504	Premature Discharge of Battery	26/03/2020	26/03/2020
							F1905	Device Revision or Replacement	A0705	Battery Problem	26/03/2020	26/03/2020
									A070504	Premature Discharge of Battery	26/03/2020	26/03/2020
					CT3001	Implantable metal	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem	26/03/2020	26/03/2020
									A070504	Premature Discharge of Battery	26/03/2020	26/03/2020
							F08	Hospitalization or Prolonged Hospitalization	A0705	Battery Problem	26/03/2020	26/03/2020
									A070504	Premature Discharge of Battery	26/03/2020	26/03/2020
							F12	Serious Injury/ Illness/ Impairment	A0705	Battery Problem	26/03/2020	26/03/2020
									A070504	Premature Discharge of Battery	26/03/2020	26/03/2020
							F19	Surgical Intervention	A0705	Battery Problem	26/03/2020	26/03/2020
									A070504	Premature Discharge of Battery	26/03/2020	26/03/2020
							F1905	Device Revision or Replacement	A0705	Battery Problem	26/03/2020	26/03/2020
									A070504	Premature Discharge of Battery	26/03/2020	26/03/2020
					CT3002	Implantable synthetic polymer	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem	26/03/2020	26/03/2020
									A070504	Premature Discharge of Battery	26/03/2020	26/03/2020
							F08	Hospitalization or Prolonged Hospitalization	A0705	Battery Problem	26/03/2020	26/03/2020
									A070504	Premature Discharge of Battery	26/03/2020	26/03/2020
							F12	Serious Injury/ Illness/ Impairment	A0705	Battery Problem	26/03/2020	26/03/2020
									A070504	Premature Discharge of Battery	26/03/2020	26/03/2020
							F19	Surgical Intervention	A0705	Battery Problem	26/03/2020	26/03/2020
									A070504	Premature Discharge of Battery	26/03/2020	26/03/2020
							F1905	Device Revision or Replacement	A0705	Battery Problem	26/03/2020	26/03/2020
									A070504	Premature Discharge of Battery	26/03/2020	26/03/2020
					CT3006	Implantable inorganic material	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem	26/03/2020	26/03/2020
									A070504	Premature Discharge of Battery	26/03/2020	26/03/2020
							F08	Hospitalization or Prolonged Hospitalization	A0705	Battery Problem	26/03/2020	26/03/2020
									A070504	Premature Discharge of Battery	26/03/2020	26/03/2020
							F12	Serious Injury/ Illness/ Impairment	A0705	Battery Problem	26/03/2020	26/03/2020
									A070504	Premature Discharge of Battery	26/03/2020	26/03/2020
							F19	Surgical Intervention	A0705	Battery Problem	26/03/2020	26/03/2020
									A070504	Premature Discharge of Battery	26/03/2020	26/03/2020
							F1905	Device Revision or Replacement	A0705	Battery Problem	26/03/2020	26/03/2020
									A070504	Premature Discharge of Battery	26/03/2020	26/03/2020
					CT314	Electrically powered	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem	26/03/2020	26/03/2020
									A070504	Premature Discharge of Battery	26/03/2020	26/03/2020
							F08	Hospitalization or Prolonged Hospitalization	A0705	Battery Problem	26/03/2020	26/03/2020
									A070504	Premature Discharge of Battery	26/03/2020	26/03/2020
							F12	Serious Injury/ Illness/ Impairment	A0705	Battery Problem	26/03/2020	26/03/2020
									A070504	Premature Discharge of Battery	26/03/2020	26/03/2020
							F19	Surgical Intervention	A0705	Battery Problem	26/03/2020	26/03/2020
									A070504	Premature Discharge of Battery	26/03/2020	26/03/2020
							F1905	Device Revision or Replacement	A0705	Battery Problem	26/03/2020	26/03/2020
									A070504	Premature Discharge of Battery	26/03/2020	26/03/2020
					CT321	Long-term surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem	26/03/2020	26/03/2020
									A070504	Premature Discharge of Battery	26/03/2020	26/03/2020
							F08	Hospitalization or Prolonged Hospitalization	A0705	Battery Problem	26/03/2020	26/03/2020
									A070504	Premature Discharge of Battery	26/03/2020	26/03/2020
							F12	Serious Injury/ Illness/ Impairment	A0705	Battery Problem	26/03/2020	26/03/2020
									A070504	Premature Discharge of Battery	26/03/2020	26/03/2020
							F19	Surgical Intervention	A0705	Battery Problem	26/03/2020	26/03/2020
									A070504	Premature Discharge of Battery	26/03/2020	26/03/2020
							F1905	Device Revision or Replacement	A0705	Battery Problem	26/03/2020	26/03/2020
									A070504	Premature Discharge of Battery	26/03/2020	26/03/2020
					CT449	Generators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem	26/03/2020	26/03/2020
									A070504	Premature Discharge of Battery	26/03/2020	26/03/2020
							F08	Hospitalization or Prolonged Hospitalization	A0705	Battery Problem	26/03/2020	26/03/2020
									A070504	Premature Discharge of Battery	26/03/2020	26/03/2020
							F12	Serious Injury/ Illness/ Impairment	A0705	Battery Problem	26/03/2020	26/03/2020
									A070504	Premature Discharge of Battery	26/03/2020	26/03/2020
							F19	Surgical Intervention	A0705	Battery Problem	26/03/2020	26/03/2020
									A070504	Premature Discharge of Battery	26/03/2020	26/03/2020
							F1905	Device Revision or Replacement	A0705	Battery Problem	26/03/2020	26/03/2020
									A070504	Premature Discharge of Battery	26/03/2020	26/03/2020
					CT981	Single use	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem	26/03/2020	26/03/2020
									A070504	Premature Discharge of Battery	26/03/2020	26/03/2020
							F08	Hospitalization or Prolonged Hospitalization	A0705	Battery Problem	26/03/2020	26/03/2020
									A070504	Premature Discharge of Battery	26/03/2020	26/03/2020
							F12	Serious Injury/ Illness/ Impairment	A0705	Battery Problem	26/03/2020	26/03/2020
									A070504	Premature Discharge of Battery	26/03/2020	26/03/2020
							F19	Surgical Intervention	A0705	Battery Problem	26/03/2020	26/03/2020
									A070504	Premature Discharge of Battery	26/03/2020	26/03/2020
							F1905	Device Revision or Replacement	A0705	Battery Problem	26/03/2020	26/03/2020
									A070504	Premature Discharge of Battery	26/03/2020	26/03/2020
					CT983	Surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem	26/03/2020	26/03/2020
									A070504	Premature Discharge of Battery	26/03/2020	26/03/2020
							F08	Hospitalization or Prolonged Hospitalization	A0705	Battery Problem	26/03/2020	26/03/2020
									A070504	Premature Discharge of Battery	26/03/2020	26/03/2020
							F12	Serious Injury/ Illness/ Impairment	A0705	Battery Problem	26/03/2020	26/03/2020
									A070504	Premature Discharge of Battery	26/03/2020	26/03/2020
							F19	Surgical Intervention	A0705	Battery Problem	26/03/2020	26/03/2020
									A070504	Premature Discharge of Battery	26/03/2020	26/03/2020
							F1905	Device Revision or Replacement	A0705	Battery Problem	26/03/2020	26/03/2020
									A070504	Premature Discharge of Battery	26/03/2020	26/03/2020
8883	2020	All other reportable incidents	37265	Dual-chamber implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		

8884	2019	All other reportable incidents	47270	Cardiac resynchronization therapy implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	F1905	Device Revision or Replacement	A0705	Battery Problem	08/05/2019	
8885	2019	All other reportable incidents	47263	Cardiac resynchronization therapy implantable pacemaker	CT1684	Cardiac pacemakers and associated devices	F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A0709	Device Sensing Problem		
					CT1685	Cardiac pacemakers	F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A0709	Device Sensing Problem		
					CT257	Implantable cardiac pacemakers	F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A0709	Device Sensing Problem		
					CT2866	Implantable cardiac electrical stimulators	F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A0709	Device Sensing Problem		
8886	2019	All other reportable incidents	17882	Physiologic-monitoring defibrillation system	CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A0709	Device Sensing Problem		
8887	2019	All other reportable incidents	17882	Physiologic-monitoring defibrillation system	CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A0203	Defective Device		
8888	2020	All other reportable incidents	47265	Dual-chamber implantable pacemaker, rate-responsive	CT141	Electrodes/Leads and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070101	Failure to Capture	26/02/2020	26/02/2020
									A071205	Pacemaker Found in Back-Up Mode	26/02/2020	26/02/2020
									A072202	Low Impedance	26/02/2020	26/02/2020
							F1901	Additional Surgery	A070101	Failure to Capture	26/02/2020	26/02/2020
									A071205	Pacemaker Found in Back-Up Mode	26/02/2020	26/02/2020
									A072202	Low Impedance	26/02/2020	26/02/2020
							F1905	Device Revision or Replacement	A070101	Failure to Capture	26/02/2020	26/02/2020
									A071205	Pacemaker Found in Back-Up Mode	26/02/2020	26/02/2020
									A072202	Low Impedance	26/02/2020	26/02/2020
					CT145	Cardiovascular devices	E2403	No Clinical Signs, Symptoms or Conditions	A070101	Failure to Capture	26/02/2020	26/02/2020
									A071205	Pacemaker Found in Back-Up Mode	26/02/2020	26/02/2020
									A072202	Low Impedance	26/02/2020	26/02/2020
							F1901	Additional Surgery	A070101	Failure to Capture	26/02/2020	26/02/2020
									A071205	Pacemaker Found in Back-Up Mode	26/02/2020	26/02/2020
									A072202	Low Impedance	26/02/2020	26/02/2020
							F1905	Device Revision or Replacement	A070101	Failure to Capture	26/02/2020	26/02/2020
									A071205	Pacemaker Found in Back-Up Mode	26/02/2020	26/02/2020
									A072202	Low Impedance	26/02/2020	26/02/2020
					CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070101	Failure to Capture	26/02/2020	26/02/2020
									A071205	Pacemaker Found in Back-Up Mode	26/02/2020	26/02/2020
									A072202	Low Impedance	26/02/2020	26/02/2020
							F1901	Additional Surgery	A070101	Failure to Capture	26/02/2020	26/02/2020
									A071205	Pacemaker Found in Back-Up Mode	26/02/2020	26/02/2020
									A072202	Low Impedance	26/02/2020	26/02/2020
							F1905	Device Revision or Replacement	A070101	Failure to Capture	26/02/2020	26/02/2020
									A071205	Pacemaker Found in Back-Up Mode	26/02/2020	26/02/2020
									A072202	Low Impedance	26/02/2020	26/02/2020
					CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A070101	Failure to Capture	26/02/2020	26/02/2020
									A071205	Pacemaker Found in Back-Up Mode	26/02/2020	26/02/2020
									A072202	Low Impedance	26/02/2020	26/02/2020
							F1901	Additional Surgery	A070101	Failure to Capture	26/02/2020	26/02/2020
									A071205	Pacemaker Found in Back-Up Mode	26/02/2020	26/02/2020
									A072202	Low Impedance	26/02/2020	26/02/2020
							F1905	Device Revision or Replacement	A070101	Failure to Capture	26/02/2020	26/02/2020
									A071205	Pacemaker Found in Back-Up Mode	26/02/2020	26/02/2020
									A072202	Low Impedance	26/02/2020	26/02/2020
					CT233	Surgical	E2403	No Clinical Signs, Symptoms or Conditions	A070101	Failure to Capture	26/02/2020	26/02/2020
									A071205	Pacemaker Found in Back-Up Mode	26/02/2020	26/02/2020
									A072202	Low Impedance	26/02/2020	26/02/2020
							F1901	Additional Surgery	A070101	Failure to Capture	26/02/2020	26/02/2020
									A071205	Pacemaker Found in Back-Up Mode	26/02/2020	26/02/2020
									A072202	Low Impedance	26/02/2020	26/02/2020
							F1905	Device Revision or Replacement	A070101	Failure to Capture	26/02/2020	26/02/2020
									A071205	Pacemaker Found in Back-Up Mode	26/02/2020	26/02/2020
									A072202	Low Impedance	26/02/2020	26/02/2020
					CT2406	Implantable	E2403	No Clinical Signs, Symptoms or Conditions	A070101	Failure to Capture	26/02/2020	26/02/2020
									A071205	Pacemaker Found in Back-Up Mode	26/02/2020	26/02/2020
									A072202	Low Impedance	26/02/2020	26/02/2020
							F1901	Additional Surgery	A070101	Failure to Capture	26/02/2020	26/02/2020
									A071205	Pacemaker Found in Back-Up Mode	26/02/2020	26/02/2020
									A072202	Low Impedance	26/02/2020	26/02/2020
							F1905	Device Revision or Replacement	A070101	Failure to Capture	26/02/2020	26/02/2020
									A071205	Pacemaker Found in Back-Up Mode	26/02/2020	26/02/2020
									A072202	Low Impedance	26/02/2020	26/02/2020
					CT2407	Active implantable	E2403	No Clinical Signs, Symptoms or Conditions	A070101	Failure to Capture	26/02/2020	26/02/2020
									A071205	Pacemaker Found in Back-Up Mode	26/02/2020	26/02/2020
									A072202	Low Impedance	26/02/2020	26/02/2020
							F1901	Additional Surgery	A070101	Failure to Capture	26/02/2020	26/02/2020
									A071205	Pacemaker Found in Back-Up Mode	26/02/2020	26/02/2020
									A072202	Low Impedance	26/02/2020	26/02/2020
							F1905	Device Revision or Replacement	A070101	Failure to Capture	26/02/2020	26/02/2020
									A071205	Pacemaker Found in Back-Up Mode	26/02/2020	26/02/2020
									A072202	Low Impedance	26/02/2020	26/02/2020
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A070101	Failure to Capture	26/02/2020	26/02/2020
									A071205	Pacemaker Found in Back-Up Mode	26/02/2020	26/02/2020
									A072202	Low Impedance	26/02/2020	26/02/2020
							F1901	Additional Surgery	A070101	Failure to Capture	26/02/2020	26/02/2020
									A071205	Pacemaker Found in Back-Up Mode	26/02/2020	26/02/2020
									A072202	Low Impedance	26/02/2020	26/02/2020
							F1905	Device Revision or Replacement	A070101	Failure to Capture	26/02/2020	26/02/2020
									A071205	Pacemaker Found in Back-Up Mode	26/02/2020	26/02/2020
									A072202	Low Impedance	26/02/2020	26/02/2020
					CT2727	Active implant devices/ancillaries	E2403	No Clinical Signs, Symptoms or Conditions	A070101	Failure to Capture	26/02/2020	26/02/2020
									A071205	Pacemaker Found in Back-Up Mode	26/02/2020	26/02/2020
									A072202	Low Impedance	26/02/2020	26/02/2020
							F1901	Additional Surgery	A070101	Failure to Capture	26/02/2020	26/02/2020
									A071205	Pacemaker Found in Back-Up Mode	26/02/2020	26/02/2020
									A072202	Low Impedance	26/02/2020	26/02/2020
							F1905	Device Revision or Replacement	A070101	Failure to Capture	26/02/2020	26/02/2020
									A071205	Pacemaker Found in Back-Up Mode	26/02/2020	26/02/2020

								A072202	Low Impedance	26/02/2020	26/02/2020
					CT2750	Implantable cardiac electrical stimulators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070101	Failure to Capture	26/02/2020	26/02/2020
									A071205	Pacemaker Found in Back-Up Mode	26/02/2020	26/02/2020
									A072202	Low Impedance	26/02/2020	26/02/2020
							F1901	Additional Surgery	A070101	Failure to Capture	26/02/2020	26/02/2020
									A071205	Pacemaker Found in Back-Up Mode	26/02/2020	26/02/2020
									A072202	Low Impedance	26/02/2020	26/02/2020
							F1905	Device Revision or Replacement	A070101	Failure to Capture	26/02/2020	26/02/2020
									A071205	Pacemaker Found in Back-Up Mode	26/02/2020	26/02/2020
									A072202	Low Impedance	26/02/2020	26/02/2020
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A070101	Failure to Capture	26/02/2020	26/02/2020
									A071205	Pacemaker Found in Back-Up Mode	26/02/2020	26/02/2020
									A072202	Low Impedance	26/02/2020	26/02/2020
							F1901	Additional Surgery	A070101	Failure to Capture	26/02/2020	26/02/2020
									A071205	Pacemaker Found in Back-Up Mode	26/02/2020	26/02/2020
									A072202	Low Impedance	26/02/2020	26/02/2020
							F1905	Device Revision or Replacement	A070101	Failure to Capture	26/02/2020	26/02/2020
									A071205	Pacemaker Found in Back-Up Mode	26/02/2020	26/02/2020
									A072202	Low Impedance	26/02/2020	26/02/2020
					CT2985	Active	E2403	No Clinical Signs, Symptoms or Conditions	A070101	Failure to Capture	26/02/2020	26/02/2020
									A071205	Pacemaker Found in Back-Up Mode	26/02/2020	26/02/2020
									A072202	Low Impedance	26/02/2020	26/02/2020
							F1901	Additional Surgery	A070101	Failure to Capture	26/02/2020	26/02/2020
									A071205	Pacemaker Found in Back-Up Mode	26/02/2020	26/02/2020
									A072202	Low Impedance	26/02/2020	26/02/2020
							F1905	Device Revision or Replacement	A070101	Failure to Capture	26/02/2020	26/02/2020
									A071205	Pacemaker Found in Back-Up Mode	26/02/2020	26/02/2020
									A072202	Low Impedance	26/02/2020	26/02/2020
					CT3001	Implantable metal	E2403	No Clinical Signs, Symptoms or Conditions	A070101	Failure to Capture	26/02/2020	26/02/2020
									A071205	Pacemaker Found in Back-Up Mode	26/02/2020	26/02/2020
									A072202	Low Impedance	26/02/2020	26/02/2020
							F1901	Additional Surgery	A070101	Failure to Capture	26/02/2020	26/02/2020
									A071205	Pacemaker Found in Back-Up Mode	26/02/2020	26/02/2020
									A072202	Low Impedance	26/02/2020	26/02/2020
							F1905	Device Revision or Replacement	A070101	Failure to Capture	26/02/2020	26/02/2020
									A071205	Pacemaker Found in Back-Up Mode	26/02/2020	26/02/2020
									A072202	Low Impedance	26/02/2020	26/02/2020
					CT3002	Implantable synthetic polymer	E2403	No Clinical Signs, Symptoms or Conditions	A070101	Failure to Capture	26/02/2020	26/02/2020
									A071205	Pacemaker Found in Back-Up Mode	26/02/2020	26/02/2020
									A072202	Low Impedance	26/02/2020	26/02/2020
							F1901	Additional Surgery	A070101	Failure to Capture	26/02/2020	26/02/2020
									A071205	Pacemaker Found in Back-Up Mode	26/02/2020	26/02/2020
									A072202	Low Impedance	26/02/2020	26/02/2020
							F1905	Device Revision or Replacement	A070101	Failure to Capture	26/02/2020	26/02/2020
									A071205	Pacemaker Found in Back-Up Mode	26/02/2020	26/02/2020
									A072202	Low Impedance	26/02/2020	26/02/2020
					CT3006	Implantable inorganic material	E2403	No Clinical Signs, Symptoms or Conditions	A070101	Failure to Capture	26/02/2020	26/02/2020
									A071205	Pacemaker Found in Back-Up Mode	26/02/2020	26/02/2020
									A072202	Low Impedance	26/02/2020	26/02/2020
							F1901	Additional Surgery	A070101	Failure to Capture	26/02/2020	26/02/2020
									A071205	Pacemaker Found in Back-Up Mode	26/02/2020	26/02/2020
									A072202	Low Impedance	26/02/2020	26/02/2020
							F1905	Device Revision or Replacement	A070101	Failure to Capture	26/02/2020	26/02/2020
									A071205	Pacemaker Found in Back-Up Mode	26/02/2020	26/02/2020
									A072202	Low Impedance	26/02/2020	26/02/2020
					CT314	Electrically powered	E2403	No Clinical Signs, Symptoms or Conditions	A070101	Failure to Capture	26/02/2020	26/02/2020
									A071205	Pacemaker Found in Back-Up Mode	26/02/2020	26/02/2020
									A072202	Low Impedance	26/02/2020	26/02/2020
							F1901	Additional Surgery	A070101	Failure to Capture	26/02/2020	26/02/2020
									A071205	Pacemaker Found in Back-Up Mode	26/02/2020	26/02/2020
									A072202	Low Impedance	26/02/2020	26/02/2020
							F1905	Device Revision or Replacement	A070101	Failure to Capture	26/02/2020	26/02/2020
									A071205	Pacemaker Found in Back-Up Mode	26/02/2020	26/02/2020
									A072202	Low Impedance	26/02/2020	26/02/2020
					CT321	Long-term surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A070101	Failure to Capture	26/02/2020	26/02/2020
									A071205	Pacemaker Found in Back-Up Mode	26/02/2020	26/02/2020
									A072202	Low Impedance	26/02/2020	26/02/2020
							F1901	Additional Surgery	A070101	Failure to Capture	26/02/2020	26/02/2020
									A071205	Pacemaker Found in Back-Up Mode	26/02/2020	26/02/2020
									A072202	Low Impedance	26/02/2020	26/02/2020
							F1905	Device Revision or Replacement	A070101	Failure to Capture	26/02/2020	26/02/2020
									A071205	Pacemaker Found in Back-Up Mode	26/02/2020	26/02/2020
									A072202	Low Impedance	26/02/2020	26/02/2020
					CT449	Generators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070101	Failure to Capture	26/02/2020	26/02/2020
									A071205	Pacemaker Found in Back-Up Mode	26/02/2020	26/02/2020
									A072202	Low Impedance	26/02/2020	26/02/2020
							F1901	Additional Surgery	A070101	Failure to Capture	26/02/2020	26/02/2020
									A071205	Pacemaker Found in Back-Up Mode	26/02/2020	26/02/2020
									A072202	Low Impedance	26/02/2020	26/02/2020
							F1905	Device Revision or Replacement	A070101	Failure to Capture	26/02/2020	26/02/2020
									A071205	Pacemaker Found in Back-Up Mode	26/02/2020	26/02/2020
									A072202	Low Impedance	26/02/2020	26/02/2020
					CT981	Single use	E2403	No Clinical Signs, Symptoms or Conditions	A070101	Failure to Capture	26/02/2020	26/02/2020
									A071205	Pacemaker Found in Back-Up Mode	26/02/2020	26/02/2020
									A072202	Low Impedance	26/02/2020	26/02/2020
							F1901	Additional Surgery	A070101	Failure to Capture	26/02/2020	26/02/2020
									A071205	Pacemaker Found in Back-Up Mode	26/02/2020	26/02/2020
									A072202	Low Impedance	26/02/2020	26/02/2020

							F1905	Device Revision or Replacement	A070101	Failure to Capture	26/02/2020	26/02/2020
									A071205	Pacemaker Found in Back-Up Mode	26/02/2020	26/02/2020
									A072202	Low Impedance	26/02/2020	26/02/2020
					CT983	Surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A070101	Failure to Capture	26/02/2020	26/02/2020
									A071205	Pacemaker Found in Back-Up Mode	26/02/2020	26/02/2020
									A072202	Low Impedance	26/02/2020	26/02/2020
							F1901	Additional Surgery	A070101	Failure to Capture	26/02/2020	26/02/2020
									A071205	Pacemaker Found in Back-Up Mode	26/02/2020	26/02/2020
									A072202	Low Impedance	26/02/2020	26/02/2020
							F1905	Device Revision or Replacement	A070101	Failure to Capture	26/02/2020	26/02/2020
									A071205	Pacemaker Found in Back-Up Mode	26/02/2020	26/02/2020
									A072202	Low Impedance	26/02/2020	26/02/2020
8889	2020	All other reportable incidents	47270	Cardiac resynchronization therapy implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	03/03/2020	03/03/2020
							F1901	Additional Surgery	A070504	Premature Discharge of Battery	03/03/2020	03/03/2020
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	03/03/2020	03/03/2020
8890	2020	All other reportable incidents	37265	Dual-chamber implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A071301	Failure to Deliver Shock/Stimulation	07/02/2020	07/02/2020
8891	2019	All other reportable incidents	35223	Endocardial/interventricular septal pacing lead	CT1684	Cardiac pacemakers and associated devices	E0506	Hemorrhage/Blood Loss/Bleeding	A0104	Migration or Expulsion of Device		
							E2114	Perforation	A0104	Migration or Expulsion of Device		
8892	2019	Death	35223	Endocardial/interventricular septal pacing lead	CT1684	Cardiac pacemakers and associated devices	F21	Unexpected Deterioration	A0709	Device Sensing Problem		
8893	2019	Death	47265	Dual-chamber implantable pacemaker, rate-responsive	CT141	Electrodes/Leads and associated devices	E2401	Insufficient Information	A0712	Pacing Problem		
					CT145	Cardiovascular devices	E2401	Insufficient Information	A0712	Pacing Problem		
					CT1684	Cardiac pacemakers and associated devices	E2401	Insufficient Information	A0712	Pacing Problem		
					CT1685	Cardiac pacemakers	E2401	Insufficient Information	A0712	Pacing Problem		
					CT233	Surgical	E2401	Insufficient Information	A0712	Pacing Problem		
					CT2406	Implantable	E2401	Insufficient Information	A0712	Pacing Problem		
					CT2407	Active implantable	E2401	Insufficient Information	A0712	Pacing Problem		
					CT257	Implantable cardiac pacemakers	E2401	Insufficient Information	A0712	Pacing Problem		
					CT2727	Active implant devices/ancillaries	E2401	Insufficient Information	A0712	Pacing Problem		
					CT2750	Implantable cardiac electrical stimulators and associated devices	E2401	Insufficient Information	A0712	Pacing Problem		
					CT2866	Implantable cardiac electrical stimulators	E2401	Insufficient Information	A0712	Pacing Problem		
					CT2985	Active	E2401	Insufficient Information	A0712	Pacing Problem		
					CT3001	Implantable metal	E2401	Insufficient Information	A0712	Pacing Problem		
					CT3002	Implantable synthetic polymer	E2401	Insufficient Information	A0712	Pacing Problem		
					CT3006	Implantable inorganic material	E2401	Insufficient Information	A0712	Pacing Problem		
					CT314	Electrically powered	E2401	Insufficient Information	A0712	Pacing Problem		
					CT321	Long-term surgical invasive	E2401	Insufficient Information	A0712	Pacing Problem		
					CT449	Generators and associated devices	E2401	Insufficient Information	A0712	Pacing Problem		
					CT981	Single use	E2401	Insufficient Information	A0712	Pacing Problem		
					CT983	Surgical invasive	E2401	Insufficient Information	A0712	Pacing Problem		
8894	2019	All other reportable incidents	37265	Dual-chamber implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E2104	Electric Shock	A070909	Oversensing		
8895	2019	All other reportable incidents	35852	Single-chamber implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A070909	Oversensing		
8896	2019	All other reportable incidents	36241	Pacemaker/defibrillator lead	CT1684	Cardiac pacemakers and associated devices	F21	Unexpected Deterioration	A	Annex A		
8897	2019	Death	47265	Dual-chamber implantable pacemaker, rate-responsive	CT141	Electrodes/Leads and associated devices	F21	Unexpected Deterioration	A0721	Circuit Failure		
					CT145	Cardiovascular devices	F21	Unexpected Deterioration	A0721	Circuit Failure		
					CT1684	Cardiac pacemakers and associated devices	F21	Unexpected Deterioration	A0721	Circuit Failure		
					CT1685	Cardiac pacemakers	F21	Unexpected Deterioration	A0721	Circuit Failure		
					CT233	Surgical	F21	Unexpected Deterioration	A0721	Circuit Failure		
					CT2406	Implantable	F21	Unexpected Deterioration	A0721	Circuit Failure		
					CT2407	Active implantable	F21	Unexpected Deterioration	A0721	Circuit Failure		
					CT257	Implantable cardiac pacemakers	F21	Unexpected Deterioration	A0721	Circuit Failure		
					CT2727	Active implant devices/ancillaries	F21	Unexpected Deterioration	A0721	Circuit Failure		
					CT2750	Implantable cardiac electrical stimulators and associated devices	F21	Unexpected Deterioration	A0721	Circuit Failure		
					CT2866	Implantable cardiac electrical stimulators	F21	Unexpected Deterioration	A0721	Circuit Failure		
					CT2985	Active	F21	Unexpected Deterioration	A0721	Circuit Failure		
					CT3001	Implantable metal	F21	Unexpected Deterioration	A0721	Circuit Failure		
					CT3002	Implantable synthetic polymer	F21	Unexpected Deterioration	A0721	Circuit Failure		
					CT3006	Implantable inorganic material	F21	Unexpected Deterioration	A0721	Circuit Failure		
					CT314	Electrically powered	F21	Unexpected Deterioration	A0721	Circuit Failure		
					CT321	Long-term surgical invasive	F21	Unexpected Deterioration	A0721	Circuit Failure		
					CT449	Generators and associated devices	F21	Unexpected Deterioration	A0721	Circuit Failure		
					CT981	Single use	F21	Unexpected Deterioration	A0721	Circuit Failure		
					CT983	Surgical invasive	F21	Unexpected Deterioration	A0721	Circuit Failure		
8898	2020	All other reportable incidents	35852	Single-chamber implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	24/03/2020	24/03/2020
							F1901	Additional Surgery	A070504	Premature Discharge of Battery	24/03/2020	24/03/2020
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	24/03/2020	24/03/2020
8899	2019	All other reportable incidents	37265	Dual-chamber implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E2104	Electric Shock	A0709	Device Sensing Problem		
							F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A0709	Device Sensing Problem		
8900	2020	All other reportable incidents	36241	Pacemaker/defibrillator lead	CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A05	Mechanical Problem	08/01/2020	08/01/2020
8901	2019	All other reportable incidents	17882	Physiologic-monitoring defibrillation system	CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A0708	Power Problem		
8902	2019	All other reportable incidents	36241	Pacemaker/defibrillator lead	CT1684	Cardiac pacemakers and associated devices	F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A0405	Degraded		
8903	2019	All other reportable incidents	35852	Single-chamber implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A070909	Oversensing		
8904	2019	All other reportable incidents	36241	Pacemaker/defibrillator lead	CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A0510	Retraction Problem	24/04/2019	
8905	2020	All other reportable incidents	36241	Pacemaker/defibrillator lead	CT1684	Cardiac pacemakers and associated devices	E2114	Perforation	A051201	Device Dislodged or Dislocated		
							F19	Surgical Intervention	A051201	Device Dislodged or Dislocated		
8906	2020	All other reportable incidents	35852	Single-chamber implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A070909	Oversensing		
8907	2020	All other reportable incidents	47270	Cardiac resynchronization therapy implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A070909	Oversensing		
							F26	No Health Consequences or Impact	A070909	Oversensing		
8908	2020	All other reportable incidents	36241	Pacemaker/defibrillator lead	CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A072201	High Impedance		
8909	2020	All other reportable incidents	36241	Pacemaker/defibrillator lead	CT1684	Cardiac pacemakers and associated devices	F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A05	Mechanical Problem	18/02/2020	18/02/2020
									A051201	Device Dislodged or Dislocated	18/02/2020	18/02/2020
							F1905	Device Revision or Replacement	A05	Mechanical Problem	18/02/2020	18/02/2020
									A051201	Device Dislodged or Dislocated	18/02/2020	18/02/2020
8910	2020	All other reportable incidents	37265	Dual-chamber implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
8911	2020	All other reportable incidents	36241	Pacemaker/defibrillator lead	CT1684	Cardiac pacemakers and associated devices	F1905	Device Revision or Replacement	A051201	Device Dislodged or Dislocated	12/02/2020	12/02/2020
8912	2020	All other reportable incidents	36241	Pacemaker/defibrillator lead	CT1684	Cardiac pacemakers and associated devices	F1905	Device Revision or Replacement	A05	Mechanical Problem		

									A051201	Device Dislodged or Dislocated		
									A070102	High Capture Threshold		
8913	2020	All other reportable incidents	35852	Single-chamber implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem		
8914	2019	All other reportable incidents	37265	Dual-chamber implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A0709	Device Sensing Problem		
8915	2020	All other reportable incidents	35852	Single-chamber implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem		
8916	2019	All other reportable incidents	36241	Pacemaker/defibrillator lead	CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A072201	High Impedance		
8917	2019	All other reportable incidents	47263	Cardiac resynchronization therapy implantable pacemaker	CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070909	Oversensing		
					CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A070909	Oversensing		
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A070909	Oversensing		
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A070909	Oversensing		
8918	2019	All other reportable incidents	37265	Dual-chamber implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery		
8919	2019	All other reportable incidents	35852	Single-chamber implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A070909	Oversensing		
8920	2019	All other reportable incidents	35852	Single-chamber implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A071302	Inappropriate/Inadequate Shock/Stimulation		
8921	2019	All other reportable incidents	37265	Dual-chamber implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	30/04/2019	
8922	2019	All other reportable incidents	35852	Single-chamber implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A071302	Inappropriate/Inadequate Shock/Stimulation		
8923	2020	All other reportable incidents	35223	Endocardial/interventricular septal pacing lead	CT1684	Cardiac pacemakers and associated devices	E0717	Dyspnea	A070909	Oversensing		
							E2403	No Clinical Signs, Symptoms or Conditions	A070909	Oversensing		
							F11	Non-Serious Injury/ Illness/ Impairment	A070909	Oversensing		
8924	2020	All other reportable incidents	47265	Dual-chamber implantable pacemaker, rate-responsive	CT141	Electrodes/Leads and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	20/03/2020	20/03/2020
							F1905	Device Revision or Replacement	A24	No Apparent Device Problem	20/03/2020	20/03/2020
					CT145	Cardiovascular devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	20/03/2020	20/03/2020
							F1905	Device Revision or Replacement	A24	No Apparent Device Problem	20/03/2020	20/03/2020
					CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	20/03/2020	20/03/2020
							F1905	Device Revision or Replacement	A24	No Apparent Device Problem	20/03/2020	20/03/2020
					CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	20/03/2020	20/03/2020
							F1905	Device Revision or Replacement	A24	No Apparent Device Problem	20/03/2020	20/03/2020
					CT233	Surgical	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	20/03/2020	20/03/2020
							F1905	Device Revision or Replacement	A24	No Apparent Device Problem	20/03/2020	20/03/2020
					CT2406	Implantable	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	20/03/2020	20/03/2020
							F1905	Device Revision or Replacement	A24	No Apparent Device Problem	20/03/2020	20/03/2020
					CT2407	Active implantable	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	20/03/2020	20/03/2020
							F1905	Device Revision or Replacement	A24	No Apparent Device Problem	20/03/2020	20/03/2020
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	20/03/2020	20/03/2020
							F1905	Device Revision or Replacement	A24	No Apparent Device Problem	20/03/2020	20/03/2020
					CT2727	Active implant devices/ancillaries	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	20/03/2020	20/03/2020
							F1905	Device Revision or Replacement	A24	No Apparent Device Problem	20/03/2020	20/03/2020
					CT2750	Implantable cardiac electrical stimulators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	20/03/2020	20/03/2020
							F1905	Device Revision or Replacement	A24	No Apparent Device Problem	20/03/2020	20/03/2020
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	20/03/2020	20/03/2020
							F1905	Device Revision or Replacement	A24	No Apparent Device Problem	20/03/2020	20/03/2020
					CT2985	Active	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	20/03/2020	20/03/2020
							F1905	Device Revision or Replacement	A24	No Apparent Device Problem	20/03/2020	20/03/2020
					CT3001	Implantable metal	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	20/03/2020	20/03/2020
							F1905	Device Revision or Replacement	A24	No Apparent Device Problem	20/03/2020	20/03/2020
					CT3002	Implantable synthetic polymer	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	20/03/2020	20/03/2020
							F1905	Device Revision or Replacement	A24	No Apparent Device Problem	20/03/2020	20/03/2020
					CT3006	Implantable inorganic material	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	20/03/2020	20/03/2020
							F1905	Device Revision or Replacement	A24	No Apparent Device Problem	20/03/2020	20/03/2020
					CT314	Electrically powered	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	20/03/2020	20/03/2020
							F1905	Device Revision or Replacement	A24	No Apparent Device Problem	20/03/2020	20/03/2020
					CT321	Long-term surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	20/03/2020	20/03/2020
							F1905	Device Revision or Replacement	A24	No Apparent Device Problem	20/03/2020	20/03/2020
					CT449	Generators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	20/03/2020	20/03/2020
							F1905	Device Revision or Replacement	A24	No Apparent Device Problem	20/03/2020	20/03/2020
					CT981	Single use	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	20/03/2020	20/03/2020
							F1905	Device Revision or Replacement	A24	No Apparent Device Problem	20/03/2020	20/03/2020
					CT983	Surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	20/03/2020	20/03/2020
							F1905	Device Revision or Replacement	A24	No Apparent Device Problem	20/03/2020	20/03/2020
8925	2019	All other reportable incidents	47265	Dual-chamber implantable pacemaker, rate-responsive	CT141	Electrodes/Leads and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
					CT145	Cardiovascular devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
					CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
					CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
					CT233	Surgical	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
					CT2406	Implantable	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
					CT2407	Active implantable	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
					CT2727	Active implant devices/ancillaries	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
					CT2750	Implantable cardiac electrical stimulators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
					CT2985	Active	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
					CT3001	Implantable metal	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
					CT3002	Implantable synthetic polymer	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
					CT3006	Implantable inorganic material	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
					CT314	Electrically powered	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
					CT321	Long-term surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
					CT449	Generators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
					CT981	Single use	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
					CT983	Surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
8926	2019	All other reportable incidents	37265	Dual-chamber implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A072202	Low Impedance	15/05/2019	15/05/2019
							F26	No Health Consequences or Impact	A072202	Low Impedance	15/05/2019	15/05/2019
8927	2019	All other reportable incidents	35852	Single-chamber implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	07/05/2019	
8928	2019	All other reportable incidents	37265	Dual-chamber implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	01/05/2019	
8929	2019	All other reportable incidents	47265	Dual-chamber implantable pacemaker, rate-responsive	CT141	Electrodes/Leads and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
					CT145	Cardiovascular devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
					CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
					CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
					CT233	Surgical	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		

					CT2406	Implantable	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
					CT2407	Active implantable	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
					CT2727	Active implant devices/ancillaries	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
					CT2750	Implantable cardiac electrical stimulators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
					CT2985	Active	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
					CT3001	Implantable metal	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
					CT3002	Implantable synthetic polymer	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
					CT3006	Implantable inorganic material	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
					CT314	Electrically powered	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
					CT321	Long-term surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
					CT449	Generators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
					CT981	Single use	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
					CT983	Surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
8930	2019	All other reportable incidents	47263	Cardiac resynchronization therapy implantable pacemaker	CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070909	Oversensing		
					CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A070909	Oversensing		
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A070909	Oversensing		
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A070909	Oversensing		
8931	2019	All other reportable incidents	35852	Single-chamber implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A070909	Oversensing		
8932	2020	All other reportable incidents	35223	Endocardial/interventricular septal pacing lead	CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070909	Oversensing	05/03/2020	05/03/2020
							F1901	Additional Surgery	A070909	Oversensing	05/03/2020	05/03/2020
							F1905	Device Revision or Replacement	A070909	Oversensing	05/03/2020	05/03/2020
8933	2020	All other reportable incidents	35223	Endocardial/interventricular septal pacing lead	CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070909	Oversensing	05/03/2020	05/03/2020
							F1901	Additional Surgery	A070909	Oversensing	05/03/2020	05/03/2020
							F1905	Device Revision or Replacement	A070909	Oversensing	05/03/2020	05/03/2020
8934	2019	All other reportable incidents	47270	Cardiac resynchronization therapy implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	13/05/2019	
8935	2019	All other reportable incidents	35223	Endocardial/interventricular septal pacing lead	CT1684	Cardiac pacemakers and associated devices	E2007	Fall	A070909	Oversensing		
							F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A070909	Oversensing		
8936	2019	All other reportable incidents	35852	Single-chamber implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	25/04/2019	
8937	2019	All other reportable incidents	36241	Pacemaker/defibrillator lead	CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A0510	Retraction Problem	23/04/2019	
8938	2019	All other reportable incidents	47267	Single-chamber implantable pacemaker, rate-responsive	CT141	Electrodes/Leads and associated devices	F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	28/05/2019	
					CT145	Cardiovascular devices	F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	28/05/2019	
					CT1684	Cardiac pacemakers and associated devices	F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	28/05/2019	
					CT1685	Cardiac pacemakers	F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	28/05/2019	
					CT233	Surgical	F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	28/05/2019	
					CT2406	Implantable	F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	28/05/2019	
					CT2407	Active implantable	F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	28/05/2019	
					CT257	Implantable cardiac pacemakers	F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	28/05/2019	
					CT2727	Active implant devices/ancillaries	F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	28/05/2019	
					CT2750	Implantable cardiac electrical stimulators and associated devices	F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	28/05/2019	
					CT2866	Implantable cardiac electrical stimulators	F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	28/05/2019	
					CT2985	Active	F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	28/05/2019	
					CT3001	Implantable metal	F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	28/05/2019	
					CT3002	Implantable synthetic polymer	F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	28/05/2019	
					CT3006	Implantable inorganic material	F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	28/05/2019	
					CT314	Electrically powered	F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	28/05/2019	
					CT321	Long-term surgical invasive	F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	28/05/2019	
					CT449	Generators and associated devices	F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	28/05/2019	
					CT981	Single use	F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	28/05/2019	
					CT983	Surgical invasive	F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	28/05/2019	
8939	2020	All other reportable incidents	37265	Dual-chamber implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	10/03/2020	10/03/2020
							F19	Surgical Intervention	A070504	Premature Discharge of Battery	10/03/2020	10/03/2020
8940	2019	All other reportable incidents	36241	Pacemaker/defibrillator lead	CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070909	Oversensing		
8941	2020	All other reportable incidents	17882	Physiologic-monitoring defibrillation system	CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070803	Failure to Power Up		
							F27	Problem Identified During Non-Clinical Procedure	A070803	Failure to Power Up		
8942	2020	All other reportable incidents	17882	Physiologic-monitoring defibrillation system	CT1684	Cardiac pacemakers and associated devices	F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A1102	Application Program Problem		
8943	2019	All other reportable incidents	47265	Dual-chamber implantable pacemaker, rate-responsive	CT141	Electrodes/Leads and associated devices	F1905	Device Revision or Replacement	A	Annex A	22/05/2019	
					CT145	Cardiovascular devices	F1905	Device Revision or Replacement	A	Annex A	22/05/2019	
					CT1684	Cardiac pacemakers and associated devices	F1905	Device Revision or Replacement	A	Annex A	22/05/2019	
					CT1685	Cardiac pacemakers	F1905	Device Revision or Replacement	A	Annex A	22/05/2019	
					CT233	Surgical	F1905	Device Revision or Replacement	A	Annex A	22/05/2019	
					CT2406	Implantable	F1905	Device Revision or Replacement	A	Annex A	22/05/2019	
					CT2407	Active implantable	F1905	Device Revision or Replacement	A	Annex A	22/05/2019	
					CT257	Implantable cardiac pacemakers	F1905	Device Revision or Replacement	A	Annex A	22/05/2019	
					CT2727	Active implant devices/ancillaries	F1905	Device Revision or Replacement	A	Annex A	22/05/2019	
					CT2750	Implantable cardiac electrical stimulators and associated devices	F1905	Device Revision or Replacement	A	Annex A	22/05/2019	
					CT2866	Implantable cardiac electrical stimulators	F1905	Device Revision or Replacement	A	Annex A	22/05/2019	
					CT2985	Active	F1905	Device Revision or Replacement	A	Annex A	22/05/2019	
					CT3001	Implantable metal	F1905	Device Revision or Replacement	A	Annex A	22/05/2019	
					CT3002	Implantable synthetic polymer	F1905	Device Revision or Replacement	A	Annex A	22/05/2019	
					CT3006	Implantable inorganic material	F1905	Device Revision or Replacement	A	Annex A	22/05/2019	
					CT314	Electrically powered	F1905	Device Revision or Replacement	A	Annex A	22/05/2019	
					CT321	Long-term surgical invasive	F1905	Device Revision or Replacement	A	Annex A	22/05/2019	
					CT449	Generators and associated devices	F1905	Device Revision or Replacement	A	Annex A	22/05/2019	
					CT981	Single use	F1905	Device Revision or Replacement	A	Annex A	22/05/2019	
					CT983	Surgical invasive	F1905	Device Revision or Replacement	A	Annex A	22/05/2019	
8944	2019	All other reportable incidents	47265	Dual-chamber implantable pacemaker, rate-responsive	CT141	Electrodes/Leads and associated devices	F1905	Device Revision or Replacement	A0721	Circuit Failure	20/05/2019	
					CT145	Cardiovascular devices	F1905	Device Revision or Replacement	A0721	Circuit Failure	20/05/2019	
					CT1684	Cardiac pacemakers and associated devices	F1905	Device Revision or Replacement	A0721	Circuit Failure	20/05/2019	
					CT1685	Cardiac pacemakers	F1905	Device Revision or Replacement	A0721	Circuit Failure	20/05/2019	
					CT233	Surgical	F1905	Device Revision or Replacement	A0721	Circuit Failure	20/05/2019	
					CT2406	Implantable	F1905	Device Revision or Replacement	A0721	Circuit Failure	20/05/2019	
					CT2407	Active implantable	F1905	Device Revision or Replacement	A0721	Circuit Failure	20/05/2019	
					CT257	Implantable cardiac pacemakers	F1905	Device Revision or Replacement	A0721	Circuit Failure	20/05/2019	

					CT2727	Active implant devices/ancillaries	F1905	Device Revision or Replacement	A0721	Circuit Failure	20/05/2019		
					CT2750	Implantable cardiac electrical stimulators and associated devices	F1905	Device Revision or Replacement	A0721	Circuit Failure	20/05/2019		
					CT2866	Implantable cardiac electrical stimulators	F1905	Device Revision or Replacement	A0721	Circuit Failure	20/05/2019		
					CT2985	Active	F1905	Device Revision or Replacement	A0721	Circuit Failure	20/05/2019		
					CT3001	Implantable metal	F1905	Device Revision or Replacement	A0721	Circuit Failure	20/05/2019		
					CT3002	Implantable synthetic polymer	F1905	Device Revision or Replacement	A0721	Circuit Failure	20/05/2019		
					CT3006	Implantable inorganic material	F1905	Device Revision or Replacement	A0721	Circuit Failure	20/05/2019		
					CT314	Electrically powered	F1905	Device Revision or Replacement	A0721	Circuit Failure	20/05/2019		
					CT321	Long-term surgical invasive	F1905	Device Revision or Replacement	A0721	Circuit Failure	20/05/2019		
					CT449	Generators and associated devices	F1905	Device Revision or Replacement	A0721	Circuit Failure	20/05/2019		
					CT981	Single use	F1905	Device Revision or Replacement	A0721	Circuit Failure	20/05/2019		
					CT983	Surgical invasive	F1905	Device Revision or Replacement	A0721	Circuit Failure	20/05/2019		
8945	2019	All other reportable incidents	47263	Cardiac resynchronization therapy implantable pacemaker	CT1684	Cardiac pacemakers and associated devices	F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	28/05/2019		
					CT1685	Cardiac pacemakers	F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	28/05/2019		
					CT257	Implantable cardiac pacemakers	F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	28/05/2019		
					CT2866	Implantable cardiac electrical stimulators	F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	28/05/2019		
8946	2020	All other reportable incidents	47267	Single-chamber implantable pacemaker, rate-responsive	CT141	Electrodes/Leads and associated devices	F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	28/02/2020	28/02/2020	
					CT145	Cardiovascular devices	F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	28/02/2020	28/02/2020	
					CT1684	Cardiac pacemakers and associated devices	F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	28/02/2020	28/02/2020	
					CT1685	Cardiac pacemakers	F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	28/02/2020	28/02/2020	
					CT233	Surgical	F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	28/02/2020	28/02/2020	
					CT2406	Implantable	F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	28/02/2020	28/02/2020	
					CT2407	Active implantable	F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	28/02/2020	28/02/2020	
					CT257	Implantable cardiac pacemakers	F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	28/02/2020	28/02/2020	
					CT2727	Active implant devices/ancillaries	F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	28/02/2020	28/02/2020	
					CT2750	Implantable cardiac electrical stimulators and associated devices	F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	28/02/2020	28/02/2020	
					CT2866	Implantable cardiac electrical stimulators	F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	28/02/2020	28/02/2020	
					CT2985	Active	F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	28/02/2020	28/02/2020	
					CT3001	Implantable metal	F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	28/02/2020	28/02/2020	
					CT3002	Implantable synthetic polymer	F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	28/02/2020	28/02/2020	
					CT3006	Implantable inorganic material	F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	28/02/2020	28/02/2020	
					CT314	Electrically powered	F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	28/02/2020	28/02/2020	
					CT321	Long-term surgical invasive	F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	28/02/2020	28/02/2020	
					CT449	Generators and associated devices	F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	28/02/2020	28/02/2020	
					CT981	Single use	F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	28/02/2020	28/02/2020	
					CT983	Surgical invasive	F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	28/02/2020	28/02/2020	
8947	2019	All other reportable incidents	47270	Cardiac resynchronization therapy implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A0709	Device Sensing Problem			
8948	2020	All other reportable incidents	47270	Cardiac resynchronization therapy implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A070910	Undersensing			
8949	2019	All other reportable incidents	47265	Dual-chamber implantable pacemaker, rate-responsive	CT141	Electrodes/Leads and associated devices	F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery			
					CT145	Cardiovascular devices	F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery			
					CT1684	Cardiac pacemakers and associated devices	F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery			
					CT1685	Cardiac pacemakers	F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery			
					CT233	Surgical	F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery			
					CT2406	Implantable	F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery			
					CT2407	Active implantable	F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery			
					CT257	Implantable cardiac pacemakers	F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery			
					CT2727	Active implant devices/ancillaries	F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery			
					CT2750	Implantable cardiac electrical stimulators and associated devices	F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery			
					CT2866	Implantable cardiac electrical stimulators	F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery			
					CT2985	Active	F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery			
					CT3001	Implantable metal	F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery			
					CT3002	Implantable synthetic polymer	F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery			
					CT3006	Implantable inorganic material	F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery			
					CT314	Electrically powered	F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery			
					CT321	Long-term surgical invasive	F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery			
					CT449	Generators and associated devices	F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery			
					CT981	Single use	F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery			
					CT983	Surgical invasive	F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery			
8950	2019	All other reportable incidents	35852	Single-chamber implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	F21	Unexpected Deterioration	A070909	Oversensing	28/04/2019		
8951	2019	All other reportable incidents	35852	Single-chamber implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A070909	Oversensing			
8952	2019	All other reportable incidents	36241	Pacemaker/defibrillator lead	CT1684	Cardiac pacemakers and associated devices	F1905	Device Revision or Replacement	A010402	Migration	29/04/2019		
8953	2020	All other reportable incidents	47270	Cardiac resynchronization therapy implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	11/03/2020	11/03/2020	
							F1901	Additional Surgery	A070504	Premature Discharge of Battery	11/03/2020	11/03/2020	
8954	2019	All other reportable incidents	35223	Endocardial/interventricular septal pacing lead	CT1684	Cardiac pacemakers and associated devices	F1905	Device Revision or Replacement	A072201	High Impedance			
8955	2020	All other reportable incidents	47270	Cardiac resynchronization therapy implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	02/03/2020	02/03/2020	
							F1901	Additional Surgery	A070504	Premature Discharge of Battery	02/03/2020	02/03/2020	
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	02/03/2020	02/03/2020	
8956	2020	All other reportable incidents	47265	Dual-chamber implantable pacemaker, rate-responsive	CT141	Electrodes/Leads and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A090801	Signal Artifact/Noise	09/01/2019	09/01/2019	
							F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A090801	Signal Artifact/Noise	09/01/2019	09/01/2019	
							F1901	Additional Surgery	A090801	Signal Artifact/Noise	09/01/2019	09/01/2019	
							F1905	Device Revision or Replacement	A090801	Signal Artifact/Noise	09/01/2019	09/01/2019	
					CT145	Cardiovascular devices	E2403	No Clinical Signs, Symptoms or Conditions	A090801	Signal Artifact/Noise	09/01/2019	09/01/2019	
							F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A090801	Signal Artifact/Noise	09/01/2019	09/01/2019	
							F1901	Additional Surgery	A090801	Signal Artifact/Noise	09/01/2019	09/01/2019	
							F1905	Device Revision or Replacement	A090801	Signal Artifact/Noise	09/01/2019	09/01/2019	
					CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A090801	Signal Artifact/Noise	09/01/2019	09/01/2019	
							F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A090801	Signal Artifact/Noise	09/01/2019	09/01/2019	
							F1901	Additional Surgery	A090801	Signal Artifact/Noise	09/01/2019	09/01/2019	
							F1905	Device Revision or Replacement	A090801	Signal Artifact/Noise	09/01/2019	09/01/2019	
					CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A090801	Signal Artifact/Noise	09/01/2019	09/01/2019	
							F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A090801	Signal Artifact/Noise	09/01/2019	09/01/2019	
							F1901	Additional Surgery	A090801	Signal Artifact/Noise	09/01/2019	09/01/2019	
							F1905	Device Revision or Replacement	A090801	Signal Artifact/Noise	09/01/2019	09/01/2019	

					CT233	Surgical	E2403	No Clinical Signs, Symptoms or Conditions	A090801	Signal Artifact/Noise	09/01/2019	09/01/2019
							F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A090801	Signal Artifact/Noise	09/01/2019	09/01/2019
							F1901	Additional Surgery	A090801	Signal Artifact/Noise	09/01/2019	09/01/2019
							F1905	Device Revision or Replacement	A090801	Signal Artifact/Noise	09/01/2019	09/01/2019
					CT2406	Implantable	E2403	No Clinical Signs, Symptoms or Conditions	A090801	Signal Artifact/Noise	09/01/2019	09/01/2019
							F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A090801	Signal Artifact/Noise	09/01/2019	09/01/2019
							F1901	Additional Surgery	A090801	Signal Artifact/Noise	09/01/2019	09/01/2019
							F1905	Device Revision or Replacement	A090801	Signal Artifact/Noise	09/01/2019	09/01/2019
					CT2407	Active implantable	E2403	No Clinical Signs, Symptoms or Conditions	A090801	Signal Artifact/Noise	09/01/2019	09/01/2019
							F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A090801	Signal Artifact/Noise	09/01/2019	09/01/2019
							F1901	Additional Surgery	A090801	Signal Artifact/Noise	09/01/2019	09/01/2019
							F1905	Device Revision or Replacement	A090801	Signal Artifact/Noise	09/01/2019	09/01/2019
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A090801	Signal Artifact/Noise	09/01/2019	09/01/2019
							F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A090801	Signal Artifact/Noise	09/01/2019	09/01/2019
							F1901	Additional Surgery	A090801	Signal Artifact/Noise	09/01/2019	09/01/2019
							F1905	Device Revision or Replacement	A090801	Signal Artifact/Noise	09/01/2019	09/01/2019
					CT2727	Active implant devices/ancillaries	E2403	No Clinical Signs, Symptoms or Conditions	A090801	Signal Artifact/Noise	09/01/2019	09/01/2019
							F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A090801	Signal Artifact/Noise	09/01/2019	09/01/2019
							F1901	Additional Surgery	A090801	Signal Artifact/Noise	09/01/2019	09/01/2019
							F1905	Device Revision or Replacement	A090801	Signal Artifact/Noise	09/01/2019	09/01/2019
					CT2750	Implantable cardiac electrical stimulators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A090801	Signal Artifact/Noise	09/01/2019	09/01/2019
							F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A090801	Signal Artifact/Noise	09/01/2019	09/01/2019
							F1901	Additional Surgery	A090801	Signal Artifact/Noise	09/01/2019	09/01/2019
							F1905	Device Revision or Replacement	A090801	Signal Artifact/Noise	09/01/2019	09/01/2019
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A090801	Signal Artifact/Noise	09/01/2019	09/01/2019
							F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A090801	Signal Artifact/Noise	09/01/2019	09/01/2019
							F1901	Additional Surgery	A090801	Signal Artifact/Noise	09/01/2019	09/01/2019
							F1905	Device Revision or Replacement	A090801	Signal Artifact/Noise	09/01/2019	09/01/2019
					CT2985	Active	E2403	No Clinical Signs, Symptoms or Conditions	A090801	Signal Artifact/Noise	09/01/2019	09/01/2019
							F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A090801	Signal Artifact/Noise	09/01/2019	09/01/2019
							F1901	Additional Surgery	A090801	Signal Artifact/Noise	09/01/2019	09/01/2019
							F1905	Device Revision or Replacement	A090801	Signal Artifact/Noise	09/01/2019	09/01/2019
					CT3001	Implantable metal	E2403	No Clinical Signs, Symptoms or Conditions	A090801	Signal Artifact/Noise	09/01/2019	09/01/2019
							F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A090801	Signal Artifact/Noise	09/01/2019	09/01/2019
							F1901	Additional Surgery	A090801	Signal Artifact/Noise	09/01/2019	09/01/2019
							F1905	Device Revision or Replacement	A090801	Signal Artifact/Noise	09/01/2019	09/01/2019
					CT3002	Implantable synthetic polymer	E2403	No Clinical Signs, Symptoms or Conditions	A090801	Signal Artifact/Noise	09/01/2019	09/01/2019
							F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A090801	Signal Artifact/Noise	09/01/2019	09/01/2019
							F1901	Additional Surgery	A090801	Signal Artifact/Noise	09/01/2019	09/01/2019
							F1905	Device Revision or Replacement	A090801	Signal Artifact/Noise	09/01/2019	09/01/2019
					CT3006	Implantable inorganic material	E2403	No Clinical Signs, Symptoms or Conditions	A090801	Signal Artifact/Noise	09/01/2019	09/01/2019
							F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A090801	Signal Artifact/Noise	09/01/2019	09/01/2019
							F1901	Additional Surgery	A090801	Signal Artifact/Noise	09/01/2019	09/01/2019
							F1905	Device Revision or Replacement	A090801	Signal Artifact/Noise	09/01/2019	09/01/2019
					CT314	Electrically powered	E2403	No Clinical Signs, Symptoms or Conditions	A090801	Signal Artifact/Noise	09/01/2019	09/01/2019
							F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A090801	Signal Artifact/Noise	09/01/2019	09/01/2019
							F1901	Additional Surgery	A090801	Signal Artifact/Noise	09/01/2019	09/01/2019
							F1905	Device Revision or Replacement	A090801	Signal Artifact/Noise	09/01/2019	09/01/2019
					CT321	Long-term surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A090801	Signal Artifact/Noise	09/01/2019	09/01/2019
							F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A090801	Signal Artifact/Noise	09/01/2019	09/01/2019
							F1901	Additional Surgery	A090801	Signal Artifact/Noise	09/01/2019	09/01/2019
							F1905	Device Revision or Replacement	A090801	Signal Artifact/Noise	09/01/2019	09/01/2019
					CT449	Generators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A090801	Signal Artifact/Noise	09/01/2019	09/01/2019
							F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A090801	Signal Artifact/Noise	09/01/2019	09/01/2019
							F1901	Additional Surgery	A090801	Signal Artifact/Noise	09/01/2019	09/01/2019
							F1905	Device Revision or Replacement	A090801	Signal Artifact/Noise	09/01/2019	09/01/2019
					CT981	Single use	E2403	No Clinical Signs, Symptoms or Conditions	A090801	Signal Artifact/Noise	09/01/2019	09/01/2019
							F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A090801	Signal Artifact/Noise	09/01/2019	09/01/2019
							F1901	Additional Surgery	A090801	Signal Artifact/Noise	09/01/2019	09/01/2019
							F1905	Device Revision or Replacement	A090801	Signal Artifact/Noise	09/01/2019	09/01/2019
					CT983	Surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A090801	Signal Artifact/Noise	09/01/2019	09/01/2019
							F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A090801	Signal Artifact/Noise	09/01/2019	09/01/2019
							F1901	Additional Surgery	A090801	Signal Artifact/Noise	09/01/2019	09/01/2019
							F1905	Device Revision or Replacement	A090801	Signal Artifact/Noise	09/01/2019	09/01/2019
8957	2020	All other reportable incidents	47265	Dual-chamber implantable pacemaker, rate-responsive	CT141	Electrodes/Leads and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A090801	Signal Artifact/Noise	09/01/2019	09/01/2019
							F1901	Additional Surgery	A090801	Signal Artifact/Noise	09/01/2019	09/01/2019
							F1905	Device Revision or Replacement	A090801	Signal Artifact/Noise	09/01/2019	09/01/2019
					CT145	Cardiovascular devices	E2403	No Clinical Signs, Symptoms or Conditions	A090801	Signal Artifact/Noise	09/01/2019	09/01/2019
							F1901	Additional Surgery	A090801	Signal Artifact/Noise	09/01/2019	09/01/2019
							F1905	Device Revision or Replacement	A090801	Signal Artifact/Noise	09/01/2019	09/01/2019
					CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A090801	Signal Artifact/Noise	09/01/2019	09/01/2019
							F1901	Additional Surgery	A090801	Signal Artifact/Noise	09/01/2019	09/01/2019
							F1905	Device Revision or Replacement	A090801	Signal Artifact/Noise	09/01/2019	09/01/2019
					CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A090801	Signal Artifact/Noise	09/01/2019	09/01/2019
							F1901	Additional Surgery	A090801	Signal Artifact/Noise	09/01/2019	09/01/2019
							F1905	Device Revision or Replacement	A090801	Signal Artifact/Noise	09/01/2019	09/01/2019
					CT233	Surgical	E2403	No Clinical Signs, Symptoms or Conditions	A090801	Signal Artifact/Noise	09/01/2019	09/01/2019
							F1901	Additional Surgery	A090801	Signal Artifact/Noise	09/01/2019	09/01/2019
							F1905	Device Revision or Replacement	A090801	Signal Artifact/Noise	09/01/2019	09/01/2019
					CT2406	Implantable	E2403	No Clinical Signs, Symptoms or Conditions	A090801	Signal Artifact/Noise	09/01/2019	09/01/2019
							F1901	Additional Surgery	A090801	Signal Artifact/Noise	09/01/2019	09/01/2019
							F1905	Device Revision or Replacement	A090801	Signal Artifact/Noise	09/01/2019	09/01/2019
					CT2407	Active implantable	E2403	No Clinical Signs, Symptoms or Conditions	A090801	Signal Artifact/Noise	09/01/2019	09/01/2019
							F1901	Additional Surgery	A090801	Signal Artifact/Noise	09/01/2019	09/01/2019
							F1905	Device Revision or Replacement	A090801	Signal Artifact/Noise	09/01/2019	09/01/2019
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A090801	Signal Artifact/Noise	09/01/2019	09/01/2019
							F1901	Additional Surgery	A090801	Signal Artifact/Noise	09/01/2019	09/01/2019
							F1905	Device Revision or Replacement	A090801	Signal Artifact/Noise	09/01/2019	09/01/2019

					CT2727	Active implant devices/ancillaries	E2403	No Clinical Signs, Symptoms or Conditions	A090801	Signal Artifact/Noise	09/01/2019	09/01/2019
							F1901	Additional Surgery	A090801	Signal Artifact/Noise	09/01/2019	09/01/2019
							F1905	Device Revision or Replacement	A090801	Signal Artifact/Noise	09/01/2019	09/01/2019
					CT2750	Implantable cardiac electrical stimulators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A090801	Signal Artifact/Noise	09/01/2019	09/01/2019
							F1901	Additional Surgery	A090801	Signal Artifact/Noise	09/01/2019	09/01/2019
							F1905	Device Revision or Replacement	A090801	Signal Artifact/Noise	09/01/2019	09/01/2019
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A090801	Signal Artifact/Noise	09/01/2019	09/01/2019
							F1901	Additional Surgery	A090801	Signal Artifact/Noise	09/01/2019	09/01/2019
							F1905	Device Revision or Replacement	A090801	Signal Artifact/Noise	09/01/2019	09/01/2019
					CT2985	Active	E2403	No Clinical Signs, Symptoms or Conditions	A090801	Signal Artifact/Noise	09/01/2019	09/01/2019
							F1901	Additional Surgery	A090801	Signal Artifact/Noise	09/01/2019	09/01/2019
							F1905	Device Revision or Replacement	A090801	Signal Artifact/Noise	09/01/2019	09/01/2019
					CT3001	Implantable metal	E2403	No Clinical Signs, Symptoms or Conditions	A090801	Signal Artifact/Noise	09/01/2019	09/01/2019
							F1901	Additional Surgery	A090801	Signal Artifact/Noise	09/01/2019	09/01/2019
							F1905	Device Revision or Replacement	A090801	Signal Artifact/Noise	09/01/2019	09/01/2019
					CT3002	Implantable synthetic polymer	E2403	No Clinical Signs, Symptoms or Conditions	A090801	Signal Artifact/Noise	09/01/2019	09/01/2019
							F1901	Additional Surgery	A090801	Signal Artifact/Noise	09/01/2019	09/01/2019
							F1905	Device Revision or Replacement	A090801	Signal Artifact/Noise	09/01/2019	09/01/2019
					CT3006	Implantable inorganic material	E2403	No Clinical Signs, Symptoms or Conditions	A090801	Signal Artifact/Noise	09/01/2019	09/01/2019
							F1901	Additional Surgery	A090801	Signal Artifact/Noise	09/01/2019	09/01/2019
							F1905	Device Revision or Replacement	A090801	Signal Artifact/Noise	09/01/2019	09/01/2019
					CT314	Electrically powered	E2403	No Clinical Signs, Symptoms or Conditions	A090801	Signal Artifact/Noise	09/01/2019	09/01/2019
							F1901	Additional Surgery	A090801	Signal Artifact/Noise	09/01/2019	09/01/2019
							F1905	Device Revision or Replacement	A090801	Signal Artifact/Noise	09/01/2019	09/01/2019
					CT321	Long-term surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A090801	Signal Artifact/Noise	09/01/2019	09/01/2019
							F1901	Additional Surgery	A090801	Signal Artifact/Noise	09/01/2019	09/01/2019
							F1905	Device Revision or Replacement	A090801	Signal Artifact/Noise	09/01/2019	09/01/2019
					CT449	Generators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A090801	Signal Artifact/Noise	09/01/2019	09/01/2019
							F1901	Additional Surgery	A090801	Signal Artifact/Noise	09/01/2019	09/01/2019
							F1905	Device Revision or Replacement	A090801	Signal Artifact/Noise	09/01/2019	09/01/2019
					CT981	Single use	E2403	No Clinical Signs, Symptoms or Conditions	A090801	Signal Artifact/Noise	09/01/2019	09/01/2019
							F1901	Additional Surgery	A090801	Signal Artifact/Noise	09/01/2019	09/01/2019
							F1905	Device Revision or Replacement	A090801	Signal Artifact/Noise	09/01/2019	09/01/2019
					CT983	Surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A090801	Signal Artifact/Noise	09/01/2019	09/01/2019
							F1901	Additional Surgery	A090801	Signal Artifact/Noise	09/01/2019	09/01/2019
							F1905	Device Revision or Replacement	A090801	Signal Artifact/Noise	09/01/2019	09/01/2019
8958	2020	All other reportable incidents	35852	Single-chamber implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	13/03/2020	13/03/2020
							F1901	Additional Surgery	A070504	Premature Discharge of Battery	13/03/2020	13/03/2020
8959	2020	All other reportable incidents	37265	Dual-chamber implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E2104	Electric Shock	A0712	Pacing Problem		
									A071302	Inappropriate/Inadequate Shock/Stimulation		
							F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A0712	Pacing Problem		
									A071302	Inappropriate/Inadequate Shock/Stimulation		
							F26	No Health Consequences or Impact	A0712	Pacing Problem		
									A071302	Inappropriate/Inadequate Shock/Stimulation		
8960	2020	All other reportable incidents	35852	Single-chamber implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	20/03/2020	20/03/2020
8961	2020	All other reportable incidents	47265	Dual-chamber implantable pacemaker, rate-responsive	CT141	Electrodes/Leads and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	19/03/2020	19/03/2020
							F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A070504	Premature Discharge of Battery	19/03/2020	19/03/2020
							F12	Serious Injury/ Illness/ Impairment	A070504	Premature Discharge of Battery	19/03/2020	19/03/2020
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	19/03/2020	19/03/2020
					CT145	Cardiovascular devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	19/03/2020	19/03/2020
							F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A070504	Premature Discharge of Battery	19/03/2020	19/03/2020
							F12	Serious Injury/ Illness/ Impairment	A070504	Premature Discharge of Battery	19/03/2020	19/03/2020
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	19/03/2020	19/03/2020
					CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	19/03/2020	19/03/2020
							F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A070504	Premature Discharge of Battery	19/03/2020	19/03/2020
							F12	Serious Injury/ Illness/ Impairment	A070504	Premature Discharge of Battery	19/03/2020	19/03/2020
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	19/03/2020	19/03/2020
					CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	19/03/2020	19/03/2020
							F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A070504	Premature Discharge of Battery	19/03/2020	19/03/2020
							F12	Serious Injury/ Illness/ Impairment	A070504	Premature Discharge of Battery	19/03/2020	19/03/2020
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	19/03/2020	19/03/2020
					CT233	Surgical	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	19/03/2020	19/03/2020
							F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A070504	Premature Discharge of Battery	19/03/2020	19/03/2020
							F12	Serious Injury/ Illness/ Impairment	A070504	Premature Discharge of Battery	19/03/2020	19/03/2020
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	19/03/2020	19/03/2020
					CT2406	Implantable	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	19/03/2020	19/03/2020
							F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A070504	Premature Discharge of Battery	19/03/2020	19/03/2020
							F12	Serious Injury/ Illness/ Impairment	A070504	Premature Discharge of Battery	19/03/2020	19/03/2020
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	19/03/2020	19/03/2020
					CT2407	Active implantable	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	19/03/2020	19/03/2020
							F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A070504	Premature Discharge of Battery	19/03/2020	19/03/2020
							F12	Serious Injury/ Illness/ Impairment	A070504	Premature Discharge of Battery	19/03/2020	19/03/2020
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	19/03/2020	19/03/2020
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	19/03/2020	19/03/2020
							F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A070504	Premature Discharge of Battery	19/03/2020	19/03/2020
							F12	Serious Injury/ Illness/ Impairment	A070504	Premature Discharge of Battery	19/03/2020	19/03/2020
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	19/03/2020	19/03/2020
					CT2727	Active implant devices/ancillaries	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	19/03/2020	19/03/2020
							F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A070504	Premature Discharge of Battery	19/03/2020	19/03/2020
							F12	Serious Injury/ Illness/ Impairment	A070504	Premature Discharge of Battery	19/03/2020	19/03/2020
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	19/03/2020	19/03/2020
					CT2750	Implantable cardiac electrical stimulators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	19/03/2020	19/03/2020
							F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A070504	Premature Discharge of Battery	19/03/2020	19/03/2020
							F12	Serious Injury/ Illness/ Impairment	A070504	Premature Discharge of Battery	19/03/2020	19/03/2020
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	19/03/2020	19/03/2020
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	19/03/2020	19/03/2020
							F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A070504	Premature Discharge of Battery	19/03/2020	19/03/2020
							F12	Serious Injury/ Illness/ Impairment	A070504	Premature Discharge of Battery	19/03/2020	19/03/2020

							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	19/03/2020	19/03/2020
					CT2985	Active	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	19/03/2020	19/03/2020
							F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A070504	Premature Discharge of Battery	19/03/2020	19/03/2020
							F12	Serious Injury/ Illness/ Impairment	A070504	Premature Discharge of Battery	19/03/2020	19/03/2020
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	19/03/2020	19/03/2020
					CT3001	Implantable metal	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	19/03/2020	19/03/2020
							F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A070504	Premature Discharge of Battery	19/03/2020	19/03/2020
							F12	Serious Injury/ Illness/ Impairment	A070504	Premature Discharge of Battery	19/03/2020	19/03/2020
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	19/03/2020	19/03/2020
					CT3002	Implantable synthetic polymer	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	19/03/2020	19/03/2020
							F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A070504	Premature Discharge of Battery	19/03/2020	19/03/2020
							F12	Serious Injury/ Illness/ Impairment	A070504	Premature Discharge of Battery	19/03/2020	19/03/2020
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	19/03/2020	19/03/2020
					CT3006	Implantable inorganic material	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	19/03/2020	19/03/2020
							F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A070504	Premature Discharge of Battery	19/03/2020	19/03/2020
							F12	Serious Injury/ Illness/ Impairment	A070504	Premature Discharge of Battery	19/03/2020	19/03/2020
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	19/03/2020	19/03/2020
					CT314	Electrically powered	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	19/03/2020	19/03/2020
							F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A070504	Premature Discharge of Battery	19/03/2020	19/03/2020
							F12	Serious Injury/ Illness/ Impairment	A070504	Premature Discharge of Battery	19/03/2020	19/03/2020
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	19/03/2020	19/03/2020
					CT321	Long-term surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	19/03/2020	19/03/2020
							F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A070504	Premature Discharge of Battery	19/03/2020	19/03/2020
							F12	Serious Injury/ Illness/ Impairment	A070504	Premature Discharge of Battery	19/03/2020	19/03/2020
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	19/03/2020	19/03/2020
					CT449	Generators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	19/03/2020	19/03/2020
							F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A070504	Premature Discharge of Battery	19/03/2020	19/03/2020
							F12	Serious Injury/ Illness/ Impairment	A070504	Premature Discharge of Battery	19/03/2020	19/03/2020
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	19/03/2020	19/03/2020
					CT981	Single use	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	19/03/2020	19/03/2020
							F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A070504	Premature Discharge of Battery	19/03/2020	19/03/2020
							F12	Serious Injury/ Illness/ Impairment	A070504	Premature Discharge of Battery	19/03/2020	19/03/2020
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	19/03/2020	19/03/2020
					CT983	Surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	19/03/2020	19/03/2020
							F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A070504	Premature Discharge of Battery	19/03/2020	19/03/2020
							F12	Serious Injury/ Illness/ Impairment	A070504	Premature Discharge of Battery	19/03/2020	19/03/2020
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	19/03/2020	19/03/2020
8962	2020	All other reportable incidents	47265	Dual-chamber implantable pacemaker, rate-responsive	CT141	Electrodes/Leads and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	01/05/2022	01/05/2022
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	01/05/2022	01/05/2022
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	01/05/2022	01/05/2022
					CT145	Cardiovascular devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	01/05/2022	01/05/2022
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	01/05/2022	01/05/2022
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	01/05/2022	01/05/2022
					CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	01/05/2022	01/05/2022
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	01/05/2022	01/05/2022
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	01/05/2022	01/05/2022
					CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	01/05/2022	01/05/2022
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	01/05/2022	01/05/2022
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	01/05/2022	01/05/2022
					CT233	Surgical	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	01/05/2022	01/05/2022
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	01/05/2022	01/05/2022
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	01/05/2022	01/05/2022
					CT2406	Implantable	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	01/05/2022	01/05/2022
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	01/05/2022	01/05/2022
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	01/05/2022	01/05/2022
					CT2407	Active implantable	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	01/05/2022	01/05/2022
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	01/05/2022	01/05/2022
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	01/05/2022	01/05/2022
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	01/05/2022	01/05/2022
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	01/05/2022	01/05/2022
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	01/05/2022	01/05/2022
					CT2727	Active implant devices/ancillaries	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	01/05/2022	01/05/2022
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	01/05/2022	01/05/2022
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	01/05/2022	01/05/2022
					CT2750	Implantable cardiac electrical stimulators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	01/05/2022	01/05/2022
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	01/05/2022	01/05/2022
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	01/05/2022	01/05/2022
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	01/05/2022	01/05/2022
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	01/05/2022	01/05/2022
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	01/05/2022	01/05/2022
					CT2985	Active	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	01/05/2022	01/05/2022
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	01/05/2022	01/05/2022
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	01/05/2022	01/05/2022
					CT3001	Implantable metal	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	01/05/2022	01/05/2022
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	01/05/2022	01/05/2022
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	01/05/2022	01/05/2022
					CT3002	Implantable synthetic polymer	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	01/05/2022	01/05/2022
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	01/05/2022	01/05/2022
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	01/05/2022	01/05/2022
					CT3006	Implantable inorganic material	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	01/05/2022	01/05/2022
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	01/05/2022	01/05/2022
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	01/05/2022	01/05/2022
					CT314	Electrically powered	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	01/05/2022	01/05/2022
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	01/05/2022	01/05/2022
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	01/05/2022	01/05/2022
					CT321	Long-term surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	01/05/2022	01/05/2022
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	01/05/2022	01/05/2022
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	01/05/2022	01/05/2022

					CT449	Generators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	01/05/2022	01/05/2022
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	01/05/2022	01/05/2022
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	01/05/2022	01/05/2022
					CT981	Single use	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	01/05/2022	01/05/2022
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	01/05/2022	01/05/2022
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	01/05/2022	01/05/2022
					CT983	Surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	01/05/2022	01/05/2022
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	01/05/2022	01/05/2022
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	01/05/2022	01/05/2022
8963	2019	All other reportable incidents	17882	Physiologic-monitoring defibrillation system	CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A09	Output Problem		
8964	2020	All other reportable incidents	36241	Pacemaker/defibrillator lead	CT1684	Cardiac pacemakers and associated devices	E1906	Unspecified Infection	A24	No Apparent Device Problem		
8965	2020	All other reportable incidents	47270	Cardiac resynchronization therapy implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E1906	Unspecified Infection	A24	No Apparent Device Problem		
8966	2020	All other reportable incidents	47267	Single-chamber implantable pacemaker, rate-responsive	CT141	Electrodes/Leads and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A090805	Incorrect Measurement	04/03/2020	04/03/2020
							F1905	Device Revision or Replacement	A090805	Incorrect Measurement	04/03/2020	04/03/2020
							F1908	Prolonged Surgery	A090805	Incorrect Measurement	04/03/2020	04/03/2020
					CT145	Cardiovascular devices	E2403	No Clinical Signs, Symptoms or Conditions	A090805	Incorrect Measurement	04/03/2020	04/03/2020
							F1905	Device Revision or Replacement	A090805	Incorrect Measurement	04/03/2020	04/03/2020
							F1908	Prolonged Surgery	A090805	Incorrect Measurement	04/03/2020	04/03/2020
					CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A090805	Incorrect Measurement	04/03/2020	04/03/2020
							F1905	Device Revision or Replacement	A090805	Incorrect Measurement	04/03/2020	04/03/2020
							F1908	Prolonged Surgery	A090805	Incorrect Measurement	04/03/2020	04/03/2020
					CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A090805	Incorrect Measurement	04/03/2020	04/03/2020
							F1905	Device Revision or Replacement	A090805	Incorrect Measurement	04/03/2020	04/03/2020
							F1908	Prolonged Surgery	A090805	Incorrect Measurement	04/03/2020	04/03/2020
					CT233	Surgical	E2403	No Clinical Signs, Symptoms or Conditions	A090805	Incorrect Measurement	04/03/2020	04/03/2020
							F1905	Device Revision or Replacement	A090805	Incorrect Measurement	04/03/2020	04/03/2020
							F1908	Prolonged Surgery	A090805	Incorrect Measurement	04/03/2020	04/03/2020
					CT2406	Implantable	E2403	No Clinical Signs, Symptoms or Conditions	A090805	Incorrect Measurement	04/03/2020	04/03/2020
							F1905	Device Revision or Replacement	A090805	Incorrect Measurement	04/03/2020	04/03/2020
							F1908	Prolonged Surgery	A090805	Incorrect Measurement	04/03/2020	04/03/2020
					CT2407	Active implantable	E2403	No Clinical Signs, Symptoms or Conditions	A090805	Incorrect Measurement	04/03/2020	04/03/2020
							F1905	Device Revision or Replacement	A090805	Incorrect Measurement	04/03/2020	04/03/2020
							F1908	Prolonged Surgery	A090805	Incorrect Measurement	04/03/2020	04/03/2020
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A090805	Incorrect Measurement	04/03/2020	04/03/2020
							F1905	Device Revision or Replacement	A090805	Incorrect Measurement	04/03/2020	04/03/2020
							F1908	Prolonged Surgery	A090805	Incorrect Measurement	04/03/2020	04/03/2020
					CT2727	Active implant devices/ancillaries	E2403	No Clinical Signs, Symptoms or Conditions	A090805	Incorrect Measurement	04/03/2020	04/03/2020
							F1905	Device Revision or Replacement	A090805	Incorrect Measurement	04/03/2020	04/03/2020
							F1908	Prolonged Surgery	A090805	Incorrect Measurement	04/03/2020	04/03/2020
					CT2750	Implantable cardiac electrical stimulators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A090805	Incorrect Measurement	04/03/2020	04/03/2020
							F1905	Device Revision or Replacement	A090805	Incorrect Measurement	04/03/2020	04/03/2020
							F1908	Prolonged Surgery	A090805	Incorrect Measurement	04/03/2020	04/03/2020
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A090805	Incorrect Measurement	04/03/2020	04/03/2020
							F1905	Device Revision or Replacement	A090805	Incorrect Measurement	04/03/2020	04/03/2020
							F1908	Prolonged Surgery	A090805	Incorrect Measurement	04/03/2020	04/03/2020
					CT2985	Active	E2403	No Clinical Signs, Symptoms or Conditions	A090805	Incorrect Measurement	04/03/2020	04/03/2020
							F1905	Device Revision or Replacement	A090805	Incorrect Measurement	04/03/2020	04/03/2020
							F1908	Prolonged Surgery	A090805	Incorrect Measurement	04/03/2020	04/03/2020
					CT3001	Implantable metal	E2403	No Clinical Signs, Symptoms or Conditions	A090805	Incorrect Measurement	04/03/2020	04/03/2020
							F1905	Device Revision or Replacement	A090805	Incorrect Measurement	04/03/2020	04/03/2020
							F1908	Prolonged Surgery	A090805	Incorrect Measurement	04/03/2020	04/03/2020
					CT3002	Implantable synthetic polymer	E2403	No Clinical Signs, Symptoms or Conditions	A090805	Incorrect Measurement	04/03/2020	04/03/2020
							F1905	Device Revision or Replacement	A090805	Incorrect Measurement	04/03/2020	04/03/2020
							F1908	Prolonged Surgery	A090805	Incorrect Measurement	04/03/2020	04/03/2020
					CT3006	Implantable inorganic material	E2403	No Clinical Signs, Symptoms or Conditions	A090805	Incorrect Measurement	04/03/2020	04/03/2020
							F1905	Device Revision or Replacement	A090805	Incorrect Measurement	04/03/2020	04/03/2020
							F1908	Prolonged Surgery	A090805	Incorrect Measurement	04/03/2020	04/03/2020
					CT314	Electrically powered	E2403	No Clinical Signs, Symptoms or Conditions	A090805	Incorrect Measurement	04/03/2020	04/03/2020
							F1905	Device Revision or Replacement	A090805	Incorrect Measurement	04/03/2020	04/03/2020
							F1908	Prolonged Surgery	A090805	Incorrect Measurement	04/03/2020	04/03/2020
					CT321	Long-term surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A090805	Incorrect Measurement	04/03/2020	04/03/2020
							F1905	Device Revision or Replacement	A090805	Incorrect Measurement	04/03/2020	04/03/2020
							F1908	Prolonged Surgery	A090805	Incorrect Measurement	04/03/2020	04/03/2020
					CT449	Generators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A090805	Incorrect Measurement	04/03/2020	04/03/2020
							F1905	Device Revision or Replacement	A090805	Incorrect Measurement	04/03/2020	04/03/2020
							F1908	Prolonged Surgery	A090805	Incorrect Measurement	04/03/2020	04/03/2020
					CT981	Single use	E2403	No Clinical Signs, Symptoms or Conditions	A090805	Incorrect Measurement	04/03/2020	04/03/2020
							F1905	Device Revision or Replacement	A090805	Incorrect Measurement	04/03/2020	04/03/2020
							F1908	Prolonged Surgery	A090805	Incorrect Measurement	04/03/2020	04/03/2020
					CT983	Surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A090805	Incorrect Measurement	04/03/2020	04/03/2020
							F1905	Device Revision or Replacement	A090805	Incorrect Measurement	04/03/2020	04/03/2020
							F1908	Prolonged Surgery	A090805	Incorrect Measurement	04/03/2020	04/03/2020
8967	2019	All other reportable incidents	47265	Dual-chamber implantable pacemaker, rate-responsive	CT141	Electrodes/Leads and associated devices	E1906	Unspecified Infection	A010102	Device Appears to Trigger Rejection		
					CT145	Cardiovascular devices	E1906	Unspecified Infection	A010102	Device Appears to Trigger Rejection		
					CT1684	Cardiac pacemakers and associated devices	E1906	Unspecified Infection	A010102	Device Appears to Trigger Rejection		
					CT1685	Cardiac pacemakers	E1906	Unspecified Infection	A010102	Device Appears to Trigger Rejection		
					CT233	Surgical	E1906	Unspecified Infection	A010102	Device Appears to Trigger Rejection		
					CT2406	Implantable	E1906	Unspecified Infection	A010102	Device Appears to Trigger Rejection		
					CT2407	Active implantable	E1906	Unspecified Infection	A010102	Device Appears to Trigger Rejection		
					CT257	Implantable cardiac pacemakers	E1906	Unspecified Infection	A010102	Device Appears to Trigger Rejection		
					CT2727	Active implant devices/ancillaries	E1906	Unspecified Infection	A010102	Device Appears to Trigger Rejection		
					CT2750	Implantable cardiac electrical stimulators and associated devices	E1906	Unspecified Infection	A010102	Device Appears to Trigger Rejection		
					CT2866	Implantable cardiac electrical stimulators	E1906	Unspecified Infection	A010102	Device Appears to Trigger Rejection		
					CT2985	Active	E1906	Unspecified Infection	A010102	Device Appears to Trigger Rejection		
					CT3001	Implantable metal	E1906	Unspecified Infection	A010102	Device Appears to Trigger Rejection		
					CT3002	Implantable synthetic polymer	E1906	Unspecified Infection	A010102	Device Appears to Trigger Rejection		

					CT3006	Implantable inorganic material	E1906	Unspecified Infection	A010102	Device Appears to Trigger Rejection		
					CT314	Electrically powered	E1906	Unspecified Infection	A010102	Device Appears to Trigger Rejection		
					CT321	Long-term surgical invasive	E1906	Unspecified Infection	A010102	Device Appears to Trigger Rejection		
					CT449	Generators and associated devices	E1906	Unspecified Infection	A010102	Device Appears to Trigger Rejection		
					CT981	Single use	E1906	Unspecified Infection	A010102	Device Appears to Trigger Rejection		
					CT983	Surgical invasive	E1906	Unspecified Infection	A010102	Device Appears to Trigger Rejection		
8968	2019	All other reportable incidents	36241	Pacemaker/defibrillator lead	CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A0722	Impedance Problem	02/05/2019	
8969	2019	All other reportable incidents	36241	Pacemaker/defibrillator lead	CT1684	Cardiac pacemakers and associated devices	F1905	Device Revision or Replacement	A072201	High Impedance	02/05/2019	
8970	2019	All other reportable incidents	47270	Cardiac resynchronization therapy implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A070909	Oversensing	28/02/2021	28/02/2021
									A072201	High Impedance	28/02/2021	28/02/2021
							F08	Hospitalization or Prolonged Hospitalization	A070909	Oversensing	28/02/2021	28/02/2021
									A072201	High Impedance	28/02/2021	28/02/2021
							F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A070909	Oversensing	28/02/2021	28/02/2021
									A072201	High Impedance	28/02/2021	28/02/2021
							F1905	Device Revision or Replacement	A070909	Oversensing	28/02/2021	28/02/2021
									A072201	High Impedance	28/02/2021	28/02/2021
8971	2019	All other reportable incidents	47267	Single-chamber implantable pacemaker, rate-responsive	CT141	Electrodes/Leads and associated devices	F1905	Device Revision or Replacement	A071205	Pacemaker Found in Back-Up Mode	01/05/2019	
					CT145	Cardiovascular devices	F1905	Device Revision or Replacement	A071205	Pacemaker Found in Back-Up Mode	01/05/2019	
					CT1684	Cardiac pacemakers and associated devices	F1905	Device Revision or Replacement	A071205	Pacemaker Found in Back-Up Mode	01/05/2019	
					CT1685	Cardiac pacemakers	F1905	Device Revision or Replacement	A071205	Pacemaker Found in Back-Up Mode	01/05/2019	
					CT233	Surgical	F1905	Device Revision or Replacement	A071205	Pacemaker Found in Back-Up Mode	01/05/2019	
					CT2406	Implantable	F1905	Device Revision or Replacement	A071205	Pacemaker Found in Back-Up Mode	01/05/2019	
					CT2407	Active implantable	F1905	Device Revision or Replacement	A071205	Pacemaker Found in Back-Up Mode	01/05/2019	
					CT257	Implantable cardiac pacemakers	F1905	Device Revision or Replacement	A071205	Pacemaker Found in Back-Up Mode	01/05/2019	
					CT2727	Active implant devices/ancillaries	F1905	Device Revision or Replacement	A071205	Pacemaker Found in Back-Up Mode	01/05/2019	
					CT2750	Implantable cardiac electrical stimulators and associated devices	F1905	Device Revision or Replacement	A071205	Pacemaker Found in Back-Up Mode	01/05/2019	
					CT2866	Implantable cardiac electrical stimulators	F1905	Device Revision or Replacement	A071205	Pacemaker Found in Back-Up Mode	01/05/2019	
					CT2985	Active	F1905	Device Revision or Replacement	A071205	Pacemaker Found in Back-Up Mode	01/05/2019	
					CT3001	Implantable metal	F1905	Device Revision or Replacement	A071205	Pacemaker Found in Back-Up Mode	01/05/2019	
					CT3002	Implantable synthetic polymer	F1905	Device Revision or Replacement	A071205	Pacemaker Found in Back-Up Mode	01/05/2019	
					CT3006	Implantable inorganic material	F1905	Device Revision or Replacement	A071205	Pacemaker Found in Back-Up Mode	01/05/2019	
					CT314	Electrically powered	F1905	Device Revision or Replacement	A071205	Pacemaker Found in Back-Up Mode	01/05/2019	
					CT321	Long-term surgical invasive	F1905	Device Revision or Replacement	A071205	Pacemaker Found in Back-Up Mode	01/05/2019	
					CT449	Generators and associated devices	F1905	Device Revision or Replacement	A071205	Pacemaker Found in Back-Up Mode	01/05/2019	
					CT981	Single use	F1905	Device Revision or Replacement	A071205	Pacemaker Found in Back-Up Mode	01/05/2019	
					CT983	Surgical invasive	F1905	Device Revision or Replacement	A071205	Pacemaker Found in Back-Up Mode	01/05/2019	
8972	2019	All other reportable incidents	47267	Single-chamber implantable pacemaker, rate-responsive	CT141	Electrodes/Leads and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A071102	Failure to Interrogate	01/05/2019	
					CT145	Cardiovascular devices	E2403	No Clinical Signs, Symptoms or Conditions	A071102	Failure to Interrogate	01/05/2019	
					CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A071102	Failure to Interrogate	01/05/2019	
					CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A071102	Failure to Interrogate	01/05/2019	
					CT233	Surgical	E2403	No Clinical Signs, Symptoms or Conditions	A071102	Failure to Interrogate	01/05/2019	
					CT2406	Implantable	E2403	No Clinical Signs, Symptoms or Conditions	A071102	Failure to Interrogate	01/05/2019	
					CT2407	Active implantable	E2403	No Clinical Signs, Symptoms or Conditions	A071102	Failure to Interrogate	01/05/2019	
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A071102	Failure to Interrogate	01/05/2019	
					CT2727	Active implant devices/ancillaries	E2403	No Clinical Signs, Symptoms or Conditions	A071102	Failure to Interrogate	01/05/2019	
					CT2750	Implantable cardiac electrical stimulators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A071102	Failure to Interrogate	01/05/2019	
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A071102	Failure to Interrogate	01/05/2019	
					CT2985	Active	E2403	No Clinical Signs, Symptoms or Conditions	A071102	Failure to Interrogate	01/05/2019	
					CT3001	Implantable metal	E2403	No Clinical Signs, Symptoms or Conditions	A071102	Failure to Interrogate	01/05/2019	
					CT3002	Implantable synthetic polymer	E2403	No Clinical Signs, Symptoms or Conditions	A071102	Failure to Interrogate	01/05/2019	
					CT3006	Implantable inorganic material	E2403	No Clinical Signs, Symptoms or Conditions	A071102	Failure to Interrogate	01/05/2019	
					CT314	Electrically powered	E2403	No Clinical Signs, Symptoms or Conditions	A071102	Failure to Interrogate	01/05/2019	
					CT321	Long-term surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A071102	Failure to Interrogate	01/05/2019	
					CT449	Generators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A071102	Failure to Interrogate	01/05/2019	
					CT981	Single use	E2403	No Clinical Signs, Symptoms or Conditions	A071102	Failure to Interrogate	01/05/2019	
					CT983	Surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A071102	Failure to Interrogate	01/05/2019	
8973	2019	All other reportable incidents	47270	Cardiac resynchronization therapy implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A150205	Difficult to Advance		
8974	2020	All other reportable incidents	37265	Dual-chamber implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A0712	Pacing Problem		
									A071302	Inappropriate/Inadequate Shock/Stimulation		
8975	2020	All other reportable incidents	37265	Dual-chamber implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	17/03/2020	14/04/2020
							F19	Surgical Intervention	A070504	Premature Discharge of Battery	17/03/2020	14/04/2020
8976	2019	All other reportable incidents	36241	Pacemaker/defibrillator lead	CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A072201	High Impedance		
8977	2019	All other reportable incidents	37265	Dual-chamber implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A072201	High Impedance		
8978	2020	All other reportable incidents	47265	Dual-chamber implantable pacemaker, rate-responsive	CT141	Electrodes/Leads and associated devices	F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A070504	Premature Discharge of Battery		
									A0712	Pacing Problem		
					CT145	Cardiovascular devices	F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A070504	Premature Discharge of Battery		
									A0712	Pacing Problem		
					CT1684	Cardiac pacemakers and associated devices	F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A070504	Premature Discharge of Battery		
									A0712	Pacing Problem		
					CT1685	Cardiac pacemakers	F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A070504	Premature Discharge of Battery		
									A0712	Pacing Problem		
					CT233	Surgical	F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A070504	Premature Discharge of Battery		
									A0712	Pacing Problem		
					CT2406	Implantable	F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A070504	Premature Discharge of Battery		
									A0712	Pacing Problem		
					CT2407	Active implantable	F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A070504	Premature Discharge of Battery		
									A0712	Pacing Problem		
					CT257	Implantable cardiac pacemakers	F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A070504	Premature Discharge of Battery		
									A0712	Pacing Problem		
					CT2727	Active implant devices/ancillaries	F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A070504	Premature Discharge of Battery		
									A0712	Pacing Problem		
					CT2750	Implantable cardiac electrical stimulators and associated devices	F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A070504	Premature Discharge of Battery		
									A0712	Pacing Problem		
					CT2866	Implantable cardiac electrical stimulators	F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A070504	Premature Discharge of Battery		
									A0712	Pacing Problem		

					CT2985	Active	F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A070504	Premature Discharge of Battery		
									A0712	Pacing Problem		
					CT3001	Implantable metal	F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A070504	Premature Discharge of Battery		
									A0712	Pacing Problem		
					CT3002	Implantable synthetic polymer	F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A070504	Premature Discharge of Battery		
									A0712	Pacing Problem		
					CT3006	Implantable inorganic material	F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A070504	Premature Discharge of Battery		
									A0712	Pacing Problem		
					CT314	Electrically powered	F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A070504	Premature Discharge of Battery		
									A0712	Pacing Problem		
					CT321	Long-term surgical invasive	F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A070504	Premature Discharge of Battery		
									A0712	Pacing Problem		
					CT449	Generators and associated devices	F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A070504	Premature Discharge of Battery		
									A0712	Pacing Problem		
					CT981	Single use	F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A070504	Premature Discharge of Battery		
									A0712	Pacing Problem		
					CT983	Surgical invasive	F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A070504	Premature Discharge of Battery		
									A0712	Pacing Problem		
8979	2020	All other reportable incidents	47265	Dual-chamber implantable pacemaker, rate-responsive	CT141	Electrodes/Leads and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	18/03/2020	18/03/2020
					CT145	Cardiovascular devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	18/03/2020	18/03/2020
					CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	18/03/2020	18/03/2020
					CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	18/03/2020	18/03/2020
					CT233	Surgical	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	18/03/2020	18/03/2020
					CT2406	Implantable	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	18/03/2020	18/03/2020
					CT2407	Active implantable	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	18/03/2020	18/03/2020
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	18/03/2020	18/03/2020
					CT2727	Active implant devices/ancillaries	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	18/03/2020	18/03/2020
					CT2750	Implantable cardiac electrical stimulators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	18/03/2020	18/03/2020
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	18/03/2020	18/03/2020
					CT2985	Active	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	18/03/2020	18/03/2020
					CT3001	Implantable metal	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	18/03/2020	18/03/2020
					CT3002	Implantable synthetic polymer	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	18/03/2020	18/03/2020
					CT3006	Implantable inorganic material	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	18/03/2020	18/03/2020
					CT314	Electrically powered	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	18/03/2020	18/03/2020
					CT321	Long-term surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	18/03/2020	18/03/2020
					CT449	Generators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	18/03/2020	18/03/2020
					CT981	Single use	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	18/03/2020	18/03/2020
					CT983	Surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	18/03/2020	18/03/2020
8980	2020	All other reportable incidents	37265	Dual-chamber implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	17/03/2020	17/03/2020
							F1901	Additional Surgery	A070504	Premature Discharge of Battery	17/03/2020	17/03/2020
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	17/03/2020	17/03/2020
8981	2019	All other reportable incidents	35852	Single-chamber implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A070909	Oversensing		
8982	2019	All other reportable incidents	36241	Pacemaker/defibrillator lead	CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070909	Oversensing		
8983	2020	All other reportable incidents	37265	Dual-chamber implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	27/03/2020	27/03/2020
									A071102	Failure to Interrogate	27/03/2020	27/03/2020
							F1901	Additional Surgery	A070504	Premature Discharge of Battery	27/03/2020	27/03/2020
									A071102	Failure to Interrogate	27/03/2020	27/03/2020
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	27/03/2020	27/03/2020
									A071102	Failure to Interrogate	27/03/2020	27/03/2020
8984	2020	All other reportable incidents	47265	Dual-chamber implantable pacemaker, rate-responsive	CT141	Electrodes/Leads and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A090804	False Positive Result		
					CT145	Cardiovascular devices	E2403	No Clinical Signs, Symptoms or Conditions	A090804	False Positive Result		
					CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A090804	False Positive Result		
					CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A090804	False Positive Result		
					CT233	Surgical	E2403	No Clinical Signs, Symptoms or Conditions	A090804	False Positive Result		
					CT2406	Implantable	E2403	No Clinical Signs, Symptoms or Conditions	A090804	False Positive Result		
					CT2407	Active implantable	E2403	No Clinical Signs, Symptoms or Conditions	A090804	False Positive Result		
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A090804	False Positive Result		
					CT2727	Active implant devices/ancillaries	E2403	No Clinical Signs, Symptoms or Conditions	A090804	False Positive Result		
					CT2750	Implantable cardiac electrical stimulators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A090804	False Positive Result		
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A090804	False Positive Result		
					CT2985	Active	E2403	No Clinical Signs, Symptoms or Conditions	A090804	False Positive Result		
					CT3001	Implantable metal	E2403	No Clinical Signs, Symptoms or Conditions	A090804	False Positive Result		
					CT3002	Implantable synthetic polymer	E2403	No Clinical Signs, Symptoms or Conditions	A090804	False Positive Result		
					CT3006	Implantable inorganic material	E2403	No Clinical Signs, Symptoms or Conditions	A090804	False Positive Result		
					CT314	Electrically powered	E2403	No Clinical Signs, Symptoms or Conditions	A090804	False Positive Result		
					CT321	Long-term surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A090804	False Positive Result		
					CT449	Generators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A090804	False Positive Result		
					CT981	Single use	E2403	No Clinical Signs, Symptoms or Conditions	A090804	False Positive Result		
					CT983	Surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A090804	False Positive Result		
8985	2020	All other reportable incidents	47270	Cardiac resynchronization therapy implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	20/03/2020	20/03/2020
							F19	Surgical Intervention	A070504	Premature Discharge of Battery	20/03/2020	20/03/2020
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	20/03/2020	20/03/2020
8986	2019	All other reportable incidents	36241	Pacemaker/defibrillator lead	CT1684	Cardiac pacemakers and associated devices	F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A010402	Migration		
							F1905	Device Revision or Replacement	A010402	Migration		
8987	2019	All other reportable incidents	47270	Cardiac resynchronization therapy implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A0512	Unintended Movement		
8988	2019	All other reportable incidents	47265	Dual-chamber implantable pacemaker, rate-responsive	CT141	Electrodes/Leads and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A072201	High Impedance		
					CT145	Cardiovascular devices	E2403	No Clinical Signs, Symptoms or Conditions	A072201	High Impedance		
					CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A072201	High Impedance		
					CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A072201	High Impedance		
					CT233	Surgical	E2403	No Clinical Signs, Symptoms or Conditions	A072201	High Impedance		
					CT2406	Implantable	E2403	No Clinical Signs, Symptoms or Conditions	A072201	High Impedance		
					CT2407	Active implantable	E2403	No Clinical Signs, Symptoms or Conditions	A072201	High Impedance		
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A072201	High Impedance		
					CT2727	Active implant devices/ancillaries	E2403	No Clinical Signs, Symptoms or Conditions	A072201	High Impedance		
					CT2750	Implantable cardiac electrical stimulators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A072201	High Impedance		

					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A072201	High Impedance		
					CT2985	Active	E2403	No Clinical Signs, Symptoms or Conditions	A072201	High Impedance		
					CT3001	Implantable metal	E2403	No Clinical Signs, Symptoms or Conditions	A072201	High Impedance		
					CT3002	Implantable synthetic polymer	E2403	No Clinical Signs, Symptoms or Conditions	A072201	High Impedance		
					CT3006	Implantable inorganic material	E2403	No Clinical Signs, Symptoms or Conditions	A072201	High Impedance		
					CT314	Electrically powered	E2403	No Clinical Signs, Symptoms or Conditions	A072201	High Impedance		
					CT321	Long-term surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A072201	High Impedance		
					CT449	Generators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A072201	High Impedance		
					CT981	Single use	E2403	No Clinical Signs, Symptoms or Conditions	A072201	High Impedance		
					CT983	Surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A072201	High Impedance		
8989	2020	All other reportable incidents	35852	Single-chamber implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A070603	Failure to Charge		
							F1905	Device Revision or Replacement	A070603	Failure to Charge		
							F26	No Health Consequences or Impact	A070603	Failure to Charge		
8990	2020	All other reportable incidents	37265	Dual-chamber implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	01/05/2020	01/05/2020
							F19	Surgical Intervention	A070504	Premature Discharge of Battery	01/05/2020	01/05/2020
8991	2020	All other reportable incidents	37265	Dual-chamber implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
8992	2020	All other reportable incidents	47265	Dual-chamber implantable pacemaker, rate-responsive	CT141	Electrodes/Leads and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070909	Oversensing		
									A0722	Impedance Problem		
									A090801	Signal Artifact/Noise		
					CT145	Cardiovascular devices	E2403	No Clinical Signs, Symptoms or Conditions	A070909	Oversensing		
									A0722	Impedance Problem		
									A090801	Signal Artifact/Noise		
					CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070909	Oversensing		
									A0722	Impedance Problem		
									A090801	Signal Artifact/Noise		
					CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A070909	Oversensing		
									A0722	Impedance Problem		
									A090801	Signal Artifact/Noise		
					CT233	Surgical	E2403	No Clinical Signs, Symptoms or Conditions	A070909	Oversensing		
									A0722	Impedance Problem		
									A090801	Signal Artifact/Noise		
					CT2406	Implantable	E2403	No Clinical Signs, Symptoms or Conditions	A070909	Oversensing		
									A0722	Impedance Problem		
									A090801	Signal Artifact/Noise		
					CT2407	Active implantable	E2403	No Clinical Signs, Symptoms or Conditions	A070909	Oversensing		
									A0722	Impedance Problem		
									A090801	Signal Artifact/Noise		
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A070909	Oversensing		
									A0722	Impedance Problem		
									A090801	Signal Artifact/Noise		
					CT2727	Active implant devices/ancillaries	E2403	No Clinical Signs, Symptoms or Conditions	A070909	Oversensing		
									A0722	Impedance Problem		
									A090801	Signal Artifact/Noise		
					CT2750	Implantable cardiac electrical stimulators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070909	Oversensing		
									A0722	Impedance Problem		
									A090801	Signal Artifact/Noise		
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A070909	Oversensing		
									A0722	Impedance Problem		
									A090801	Signal Artifact/Noise		
					CT2985	Active	E2403	No Clinical Signs, Symptoms or Conditions	A070909	Oversensing		
									A0722	Impedance Problem		
									A090801	Signal Artifact/Noise		
					CT3001	Implantable metal	E2403	No Clinical Signs, Symptoms or Conditions	A070909	Oversensing		
									A0722	Impedance Problem		
									A090801	Signal Artifact/Noise		
					CT3002	Implantable synthetic polymer	E2403	No Clinical Signs, Symptoms or Conditions	A070909	Oversensing		
									A0722	Impedance Problem		
									A090801	Signal Artifact/Noise		
					CT3006	Implantable inorganic material	E2403	No Clinical Signs, Symptoms or Conditions	A070909	Oversensing		
									A0722	Impedance Problem		
									A090801	Signal Artifact/Noise		
					CT314	Electrically powered	E2403	No Clinical Signs, Symptoms or Conditions	A070909	Oversensing		
									A0722	Impedance Problem		
									A090801	Signal Artifact/Noise		
					CT321	Long-term surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A070909	Oversensing		
									A0722	Impedance Problem		
									A090801	Signal Artifact/Noise		
					CT449	Generators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070909	Oversensing		
									A0722	Impedance Problem		
									A090801	Signal Artifact/Noise		
					CT981	Single use	E2403	No Clinical Signs, Symptoms or Conditions	A070909	Oversensing		
									A0722	Impedance Problem		
									A090801	Signal Artifact/Noise		
					CT983	Surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A070909	Oversensing		
									A0722	Impedance Problem		
									A090801	Signal Artifact/Noise		
8993	2020	All other reportable incidents	47263	Cardiac resynchronization therapy implantable pacemaker	CT1684	Cardiac pacemakers and associated devices	E0112	Dizziness	A070909	Oversensing		
							E2403	No Clinical Signs, Symptoms or Conditions	A070909	Oversensing		
							F15	Recognized Device or Procedural Complication	A070909	Oversensing		
					CT1685	Cardiac pacemakers	E0112	Dizziness	A070909	Oversensing		
							E2403	No Clinical Signs, Symptoms or Conditions	A070909	Oversensing		
							F15	Recognized Device or Procedural Complication	A070909	Oversensing		
					CT257	Implantable cardiac pacemakers	E0112	Dizziness	A070909	Oversensing		
							E2403	No Clinical Signs, Symptoms or Conditions	A070909	Oversensing		
							F15	Recognized Device or Procedural Complication	A070909	Oversensing		
					CT2866	Implantable cardiac electrical stimulators	E0112	Dizziness	A070909	Oversensing		
							E2403	No Clinical Signs, Symptoms or Conditions	A070909	Oversensing		
							F15	Recognized Device or Procedural Complication	A070909	Oversensing		

8994	2019	All other reportable incidents	17882	Physiologic-monitoring defibrillation system	CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A110201	Application Program Freezes or Fails to Launch		
8995	2020	All other reportable incidents	36241	Pacemaker/defibrillator lead	CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A072201	High Impedance		
									A090801	Signal Artifact/Noise		
							F26	No Health Consequences or Impact	A072201	High Impedance		
									A090801	Signal Artifact/Noise		
8996	2020	All other reportable incidents	36241	Pacemaker/defibrillator lead	CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A072201	High Impedance		
8997	2020	All other reportable incidents	35852	Single-chamber implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A072202	Low Impedance		
							F26	No Health Consequences or Impact	A072202	Low Impedance		
8998	2020	All other reportable incidents	35852	Single-chamber implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem	11/03/2020	11/03/2020
									A1601	Device Alarm System	11/03/2020	11/03/2020
							F08	Hospitalization or Prolonged Hospitalization	A0705	Battery Problem	11/03/2020	11/03/2020
									A1601	Device Alarm System	11/03/2020	11/03/2020
							F12	Serious Injury/ Illness/ Impairment	A0705	Battery Problem	11/03/2020	11/03/2020
									A1601	Device Alarm System	11/03/2020	11/03/2020
							F19	Surgical Intervention	A0705	Battery Problem	11/03/2020	11/03/2020
									A1601	Device Alarm System	11/03/2020	11/03/2020
							F1905	Device Revision or Replacement	A0705	Battery Problem	11/03/2020	11/03/2020
									A1601	Device Alarm System	11/03/2020	11/03/2020
8999	2020	All other reportable incidents	36241	Pacemaker/defibrillator lead	CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A040101	Fracture	06/03/2020	06/03/2020
									A05	Mechanical Problem	06/03/2020	06/03/2020
							F1905	Device Revision or Replacement	A040101	Fracture	06/03/2020	06/03/2020
									A05	Mechanical Problem	06/03/2020	06/03/2020
							F26	No Health Consequences or Impact	A040101	Fracture	06/03/2020	06/03/2020
									A05	Mechanical Problem	06/03/2020	06/03/2020
9000	2020	All other reportable incidents	47265	Dual-chamber implantable pacemaker, rate-responsive	CT141	Electrodes/Leads and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT145	Cardiovascular devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT233	Surgical	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT2406	Implantable	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT2407	Active implantable	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT2727	Active implant devices/ancillaries	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT2750	Implantable cardiac electrical stimulators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT2985	Active	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT3001	Implantable metal	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT3002	Implantable synthetic polymer	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT3006	Implantable inorganic material	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT314	Electrically powered	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT321	Long-term surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT449	Generators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT981	Single use	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT983	Surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
9001	2020	All other reportable incidents	47265	Dual-chamber implantable pacemaker, rate-responsive	CT141	Electrodes/Leads and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT145	Cardiovascular devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT233	Surgical	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT2406	Implantable	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT2407	Active implantable	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT2727	Active implant devices/ancillaries	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT2750	Implantable cardiac electrical stimulators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT2985	Active	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		

					CT3001	Implantable metal	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT3002	Implantable synthetic polymer	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT3006	Implantable inorganic material	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT314	Electrically powered	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT321	Long-term surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT449	Generators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT981	Single use	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT983	Surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
9002	2020	All other reportable incidents	47265	Dual-chamber implantable pacemaker, rate-responsive	CT141	Electrodes/Leads and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT145	Cardiovascular devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT233	Surgical	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT2406	Implantable	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT2407	Active implantable	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT2727	Active implant devices/ancillaries	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT2750	Implantable cardiac electrical stimulators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT2985	Active	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT3001	Implantable metal	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT3002	Implantable synthetic polymer	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT3006	Implantable inorganic material	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT314	Electrically powered	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT321	Long-term surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT449	Generators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT981	Single use	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT983	Surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
9003	2020	All other reportable incidents	47265	Dual-chamber implantable pacemaker, rate-responsive	CT141	Electrodes/Leads and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	24/03/2020	24/03/2020
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	24/03/2020	24/03/2020
							F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A070504	Premature Discharge of Battery	24/03/2020	24/03/2020
							F12	Serious Injury/ Illness/ Impairment	A070504	Premature Discharge of Battery	24/03/2020	24/03/2020
							F19	Surgical Intervention	A070504	Premature Discharge of Battery	24/03/2020	24/03/2020
					CT145	Cardiovascular devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	24/03/2020	24/03/2020
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	24/03/2020	24/03/2020
							F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A070504	Premature Discharge of Battery	24/03/2020	24/03/2020
							F12	Serious Injury/ Illness/ Impairment	A070504	Premature Discharge of Battery	24/03/2020	24/03/2020
							F19	Surgical Intervention	A070504	Premature Discharge of Battery	24/03/2020	24/03/2020
					CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	24/03/2020	24/03/2020
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	24/03/2020	24/03/2020
							F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A070504	Premature Discharge of Battery	24/03/2020	24/03/2020
							F12	Serious Injury/ Illness/ Impairment	A070504	Premature Discharge of Battery	24/03/2020	24/03/2020
							F19	Surgical Intervention	A070504	Premature Discharge of Battery	24/03/2020	24/03/2020
					CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	24/03/2020	24/03/2020
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	24/03/2020	24/03/2020
							F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A070504	Premature Discharge of Battery	24/03/2020	24/03/2020
							F12	Serious Injury/ Illness/ Impairment	A070504	Premature Discharge of Battery	24/03/2020	24/03/2020
							F19	Surgical Intervention	A070504	Premature Discharge of Battery	24/03/2020	24/03/2020
					CT233	Surgical	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	24/03/2020	24/03/2020
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	24/03/2020	24/03/2020
							F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A070504	Premature Discharge of Battery	24/03/2020	24/03/2020
							F12	Serious Injury/ Illness/ Impairment	A070504	Premature Discharge of Battery	24/03/2020	24/03/2020
							F19	Surgical Intervention	A070504	Premature Discharge of Battery	24/03/2020	24/03/2020
					CT2406	Implantable	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	24/03/2020	24/03/2020
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	24/03/2020	24/03/2020
							F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A070504	Premature Discharge of Battery	24/03/2020	24/03/2020
							F12	Serious Injury/ Illness/ Impairment	A070504	Premature Discharge of Battery	24/03/2020	24/03/2020
							F19	Surgical Intervention	A070504	Premature Discharge of Battery	24/03/2020	24/03/2020
					CT2407	Active implantable	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	24/03/2020	24/03/2020
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	24/03/2020	24/03/2020

							F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A070504	Premature Discharge of Battery	24/03/2020	24/03/2020
							F12	Serious Injury/ Illness/ Impairment	A070504	Premature Discharge of Battery	24/03/2020	24/03/2020
							F19	Surgical Intervention	A070504	Premature Discharge of Battery	24/03/2020	24/03/2020
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	24/03/2020	24/03/2020
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	24/03/2020	24/03/2020
							F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A070504	Premature Discharge of Battery	24/03/2020	24/03/2020
							F12	Serious Injury/ Illness/ Impairment	A070504	Premature Discharge of Battery	24/03/2020	24/03/2020
							F19	Surgical Intervention	A070504	Premature Discharge of Battery	24/03/2020	24/03/2020
					CT2727	Active implant devices/ancillaries	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	24/03/2020	24/03/2020
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	24/03/2020	24/03/2020
							F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A070504	Premature Discharge of Battery	24/03/2020	24/03/2020
							F12	Serious Injury/ Illness/ Impairment	A070504	Premature Discharge of Battery	24/03/2020	24/03/2020
							F19	Surgical Intervention	A070504	Premature Discharge of Battery	24/03/2020	24/03/2020
					CT2750	Implantable cardiac electrical stimulators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	24/03/2020	24/03/2020
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	24/03/2020	24/03/2020
							F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A070504	Premature Discharge of Battery	24/03/2020	24/03/2020
							F12	Serious Injury/ Illness/ Impairment	A070504	Premature Discharge of Battery	24/03/2020	24/03/2020
							F19	Surgical Intervention	A070504	Premature Discharge of Battery	24/03/2020	24/03/2020
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	24/03/2020	24/03/2020
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	24/03/2020	24/03/2020
							F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A070504	Premature Discharge of Battery	24/03/2020	24/03/2020
							F12	Serious Injury/ Illness/ Impairment	A070504	Premature Discharge of Battery	24/03/2020	24/03/2020
							F19	Surgical Intervention	A070504	Premature Discharge of Battery	24/03/2020	24/03/2020
					CT2985	Active	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	24/03/2020	24/03/2020
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	24/03/2020	24/03/2020
							F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A070504	Premature Discharge of Battery	24/03/2020	24/03/2020
							F12	Serious Injury/ Illness/ Impairment	A070504	Premature Discharge of Battery	24/03/2020	24/03/2020
							F19	Surgical Intervention	A070504	Premature Discharge of Battery	24/03/2020	24/03/2020
					CT3001	Implantable metal	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	24/03/2020	24/03/2020
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	24/03/2020	24/03/2020
							F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A070504	Premature Discharge of Battery	24/03/2020	24/03/2020
							F12	Serious Injury/ Illness/ Impairment	A070504	Premature Discharge of Battery	24/03/2020	24/03/2020
							F19	Surgical Intervention	A070504	Premature Discharge of Battery	24/03/2020	24/03/2020
					CT3002	Implantable synthetic polymer	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	24/03/2020	24/03/2020
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	24/03/2020	24/03/2020
							F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A070504	Premature Discharge of Battery	24/03/2020	24/03/2020
							F12	Serious Injury/ Illness/ Impairment	A070504	Premature Discharge of Battery	24/03/2020	24/03/2020
							F19	Surgical Intervention	A070504	Premature Discharge of Battery	24/03/2020	24/03/2020
					CT3006	Implantable inorganic material	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	24/03/2020	24/03/2020
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	24/03/2020	24/03/2020
							F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A070504	Premature Discharge of Battery	24/03/2020	24/03/2020
							F12	Serious Injury/ Illness/ Impairment	A070504	Premature Discharge of Battery	24/03/2020	24/03/2020
							F19	Surgical Intervention	A070504	Premature Discharge of Battery	24/03/2020	24/03/2020
					CT314	Electrically powered	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	24/03/2020	24/03/2020
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	24/03/2020	24/03/2020
							F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A070504	Premature Discharge of Battery	24/03/2020	24/03/2020
							F12	Serious Injury/ Illness/ Impairment	A070504	Premature Discharge of Battery	24/03/2020	24/03/2020
							F19	Surgical Intervention	A070504	Premature Discharge of Battery	24/03/2020	24/03/2020
					CT321	Long-term surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	24/03/2020	24/03/2020
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	24/03/2020	24/03/2020
							F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A070504	Premature Discharge of Battery	24/03/2020	24/03/2020
							F12	Serious Injury/ Illness/ Impairment	A070504	Premature Discharge of Battery	24/03/2020	24/03/2020
							F19	Surgical Intervention	A070504	Premature Discharge of Battery	24/03/2020	24/03/2020
					CT449	Generators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	24/03/2020	24/03/2020
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	24/03/2020	24/03/2020
							F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A070504	Premature Discharge of Battery	24/03/2020	24/03/2020
							F12	Serious Injury/ Illness/ Impairment	A070504	Premature Discharge of Battery	24/03/2020	24/03/2020
							F19	Surgical Intervention	A070504	Premature Discharge of Battery	24/03/2020	24/03/2020
					CT981	Single use	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	24/03/2020	24/03/2020
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	24/03/2020	24/03/2020
							F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A070504	Premature Discharge of Battery	24/03/2020	24/03/2020
							F12	Serious Injury/ Illness/ Impairment	A070504	Premature Discharge of Battery	24/03/2020	24/03/2020
							F19	Surgical Intervention	A070504	Premature Discharge of Battery	24/03/2020	24/03/2020
					CT983	Surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	24/03/2020	24/03/2020
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	24/03/2020	24/03/2020
							F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A070504	Premature Discharge of Battery	24/03/2020	24/03/2020
							F12	Serious Injury/ Illness/ Impairment	A070504	Premature Discharge of Battery	24/03/2020	24/03/2020
							F19	Surgical Intervention	A070504	Premature Discharge of Battery	24/03/2020	24/03/2020
9004	2019	All other reportable incidents	35223	Endocardial/interventricular septal pacing lead	CT1684	Cardiac pacemakers and associated devices	F1905	Device Revision or Replacement	A070909	Oversensing		
9005	2019	All other reportable incidents	47265	Dual-chamber implantable pacemaker, rate-responsive	CT141	Electrodes/Leads and associated devices	E2007	Fall	A070909	Oversensing		
							F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A070909	Oversensing		
					CT145	Cardiovascular devices	E2007	Fall	A070909	Oversensing		
							F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A070909	Oversensing		
					CT1684	Cardiac pacemakers and associated devices	E2007	Fall	A070909	Oversensing		
							F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A070909	Oversensing		
					CT1685	Cardiac pacemakers	E2007	Fall	A070909	Oversensing		
							F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A070909	Oversensing		
					CT233	Surgical	E2007	Fall	A070909	Oversensing		
							F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A070909	Oversensing		
					CT2406	Implantable	E2007	Fall	A070909	Oversensing		
							F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A070909	Oversensing		
					CT2407	Active implantable	E2007	Fall	A070909	Oversensing		
							F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A070909	Oversensing		
					CT257	Implantable cardiac pacemakers	E2007	Fall	A070909	Oversensing		
							F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A070909	Oversensing		
					CT2727	Active implant devices/ancillaries	E2007	Fall	A070909	Oversensing		
							F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A070909	Oversensing		

					CT2750	Implantable cardiac electrical stimulators and associated devices	E2007	Fall	A070909	Oversensing		
							F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A070909	Oversensing		
					CT2866	Implantable cardiac electrical stimulators	E2007	Fall	A070909	Oversensing		
							F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A070909	Oversensing		
					CT2985	Active	E2007	Fall	A070909	Oversensing		
							F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A070909	Oversensing		
					CT3001	Implantable metal	E2007	Fall	A070909	Oversensing		
							F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A070909	Oversensing		
					CT3002	Implantable synthetic polymer	E2007	Fall	A070909	Oversensing		
							F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A070909	Oversensing		
					CT3006	Implantable inorganic material	E2007	Fall	A070909	Oversensing		
							F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A070909	Oversensing		
					CT314	Electrically powered	E2007	Fall	A070909	Oversensing		
							F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A070909	Oversensing		
					CT321	Long-term surgical invasive	E2007	Fall	A070909	Oversensing		
							F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A070909	Oversensing		
					CT449	Generators and associated devices	E2007	Fall	A070909	Oversensing		
							F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A070909	Oversensing		
					CT981	Single use	E2007	Fall	A070909	Oversensing		
							F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A070909	Oversensing		
					CT983	Surgical invasive	E2007	Fall	A070909	Oversensing		
							F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A070909	Oversensing		
9006	2019	All other reportable incidents	47270	Cardiac resynchronization therapy implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	F1905	Device Revision or Replacement	A071205	Pacemaker Found in Back-Up Mode	20/05/2019	
9007	2019	All other reportable incidents	35852	Single-chamber implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A071302	Inappropriate/Inadequate Shock/Stimulation		
9008	2020	All other reportable incidents	36241	Pacemaker/defibrillator lead	CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070910	Undersensing		
									A1301	Failure to Read Input Signal		
							F15	Recognized Device or Procedural Complication	A070910	Undersensing		
									A1301	Failure to Read Input Signal		
9009	2020	All other reportable incidents	35852	Single-chamber implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A1304	Telemetry Discrepancy		
9010	2020	All other reportable incidents	47265	Dual-chamber implantable pacemaker, rate-responsive	CT141	Electrodes/Leads and associated devices	E060109	Tachycardia	A0203	Defective Device		
									A070504	Premature Discharge of Battery		
					CT145	Cardiovascular devices	E060109	Tachycardia	A0203	Defective Device		
									A070504	Premature Discharge of Battery		
					CT1684	Cardiac pacemakers and associated devices	E060109	Tachycardia	A0203	Defective Device		
									A070504	Premature Discharge of Battery		
					CT1685	Cardiac pacemakers	E060109	Tachycardia	A0203	Defective Device		
									A070504	Premature Discharge of Battery		
					CT233	Surgical	E060109	Tachycardia	A0203	Defective Device		
									A070504	Premature Discharge of Battery		
					CT2406	Implantable	E060109	Tachycardia	A0203	Defective Device		
									A070504	Premature Discharge of Battery		
					CT2407	Active implantable	E060109	Tachycardia	A0203	Defective Device		
									A070504	Premature Discharge of Battery		
					CT257	Implantable cardiac pacemakers	E060109	Tachycardia	A0203	Defective Device		
									A070504	Premature Discharge of Battery		
					CT2727	Active implant devices/ancillaries	E060109	Tachycardia	A0203	Defective Device		
									A070504	Premature Discharge of Battery		
					CT2750	Implantable cardiac electrical stimulators and associated devices	E060109	Tachycardia	A0203	Defective Device		
									A070504	Premature Discharge of Battery		
					CT2866	Implantable cardiac electrical stimulators	E060109	Tachycardia	A0203	Defective Device		
									A070504	Premature Discharge of Battery		
					CT2985	Active	E060109	Tachycardia	A0203	Defective Device		
									A070504	Premature Discharge of Battery		
					CT3001	Implantable metal	E060109	Tachycardia	A0203	Defective Device		
									A070504	Premature Discharge of Battery		
					CT3002	Implantable synthetic polymer	E060109	Tachycardia	A0203	Defective Device		
									A070504	Premature Discharge of Battery		
					CT3006	Implantable inorganic material	E060109	Tachycardia	A0203	Defective Device		
									A070504	Premature Discharge of Battery		
					CT314	Electrically powered	E060109	Tachycardia	A0203	Defective Device		
									A070504	Premature Discharge of Battery		
					CT321	Long-term surgical invasive	E060109	Tachycardia	A0203	Defective Device		
									A070504	Premature Discharge of Battery		
					CT449	Generators and associated devices	E060109	Tachycardia	A0203	Defective Device		
									A070504	Premature Discharge of Battery		
					CT981	Single use	E060109	Tachycardia	A0203	Defective Device		
									A070504	Premature Discharge of Battery		
					CT983	Surgical invasive	E060109	Tachycardia	A0203	Defective Device		
									A070504	Premature Discharge of Battery		
9011	2020	All other reportable incidents	35852	Single-chamber implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	19/03/2020	09/07/2020
							F19	Surgical Intervention	A070504	Premature Discharge of Battery	19/03/2020	09/07/2020
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	19/03/2020	09/07/2020
9012	2020	All other reportable incidents	37265	Dual-chamber implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	02/04/2020	02/04/2020
							F19	Surgical Intervention	A070504	Premature Discharge of Battery	02/04/2020	02/04/2020
9013	2019	All other reportable incidents	47265	Dual-chamber implantable pacemaker, rate-responsive	CT141	Electrodes/Leads and associated devices	F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	20/03/2019	
					CT145	Cardiovascular devices	F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	20/03/2019	
					CT1684	Cardiac pacemakers and associated devices	F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	20/03/2019	
					CT1685	Cardiac pacemakers	F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	20/03/2019	
					CT233	Surgical	F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	20/03/2019	
					CT2406	Implantable	F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	20/03/2019	
					CT2407	Active implantable	F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	20/03/2019	
					CT257	Implantable cardiac pacemakers	F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	20/03/2019	
					CT2727	Active implant devices/ancillaries	F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	20/03/2019	
					CT2750	Implantable cardiac electrical stimulators and associated devices	F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	20/03/2019	
					CT2866	Implantable cardiac electrical stimulators	F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	20/03/2019	
					CT2985	Active	F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	20/03/2019	

					CT3001	Implantable metal	F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	20/03/2019	
					CT3002	Implantable synthetic polymer	F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	20/03/2019	
					CT3006	Implantable inorganic material	F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	20/03/2019	
					CT314	Electrically powered	F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	20/03/2019	
					CT321	Long-term surgical invasive	F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	20/03/2019	
					CT449	Generators and associated devices	F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	20/03/2019	
					CT981	Single use	F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	20/03/2019	
					CT983	Surgical invasive	F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	20/03/2019	
9014	2020	All other reportable incidents	37265	Dual-chamber implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A070909	Oversensing		
							F1901	Additional Surgery	A070909	Oversensing		
							F1905	Device Revision or Replacement	A070909	Oversensing		
9015	2019	All other reportable incidents	37265	Dual-chamber implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	24/05/2019	
9016	2020	All other reportable incidents	35223	Endocardial/interventricular septal pacing lead	CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A090801	Signal Artifact/Noise	30/01/2020	30/01/2020
							F1901	Additional Surgery	A090801	Signal Artifact/Noise	30/01/2020	30/01/2020
9017	2020	Unanticipated serious deterioration in state of health	17882	Physiologic-monitoring defibrillation system	CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A07	Electrical /Electronic Property Problem		
									A0712	Pacing Problem		
									A11	Computer Software Problem		
									A1102	Application Program Problem		
									A110201	Application Program Freezes or Fails to Launch		
							F05	Delay to Treatment/ Therapy	A07	Electrical /Electronic Property Problem		
									A0712	Pacing Problem		
									A11	Computer Software Problem		
									A1102	Application Program Problem		
									A110201	Application Program Freezes or Fails to Launch		
9018	2019	All other reportable incidents	60789	Intracardiac pacemaker	CT130	Catheters and associated devices	F1905	Device Revision or Replacement	A25	No Apparent Adverse Event	03/06/2019	
					CT145	Cardiovascular devices	F1905	Device Revision or Replacement	A25	No Apparent Adverse Event	03/06/2019	
					CT1684	Cardiac pacemakers and associated devices	F1905	Device Revision or Replacement	A25	No Apparent Adverse Event	03/06/2019	
					CT1685	Cardiac pacemakers	F1905	Device Revision or Replacement	A25	No Apparent Adverse Event	03/06/2019	
					CT233	Surgical	F1905	Device Revision or Replacement	A25	No Apparent Adverse Event	03/06/2019	
					CT2406	Implantable	F1905	Device Revision or Replacement	A25	No Apparent Adverse Event	03/06/2019	
					CT2407	Active implantable	F1905	Device Revision or Replacement	A25	No Apparent Adverse Event	03/06/2019	
					CT257	Implantable cardiac pacemakers	F1905	Device Revision or Replacement	A25	No Apparent Adverse Event	03/06/2019	
					CT2727	Active implant devices/ancillaries	F1905	Device Revision or Replacement	A25	No Apparent Adverse Event	03/06/2019	
					CT2750	Implantable cardiac electrical stimulators and associated devices	F1905	Device Revision or Replacement	A25	No Apparent Adverse Event	03/06/2019	
					CT2818	Catheterization/Cannulation/Fluid path devices	F1905	Device Revision or Replacement	A25	No Apparent Adverse Event	03/06/2019	
					CT2866	Implantable cardiac electrical stimulators	F1905	Device Revision or Replacement	A25	No Apparent Adverse Event	03/06/2019	
					CT2985	Active	F1905	Device Revision or Replacement	A25	No Apparent Adverse Event	03/06/2019	
					CT3001	Implantable metal	F1905	Device Revision or Replacement	A25	No Apparent Adverse Event	03/06/2019	
					CT3002	Implantable synthetic polymer	F1905	Device Revision or Replacement	A25	No Apparent Adverse Event	03/06/2019	
					CT3006	Implantable inorganic material	F1905	Device Revision or Replacement	A25	No Apparent Adverse Event	03/06/2019	
					CT314	Electrically powered	F1905	Device Revision or Replacement	A25	No Apparent Adverse Event	03/06/2019	
					CT321	Long-term surgical invasive	F1905	Device Revision or Replacement	A25	No Apparent Adverse Event	03/06/2019	
					CT981	Single use	F1905	Device Revision or Replacement	A25	No Apparent Adverse Event	03/06/2019	
					CT983	Surgical invasive	F1905	Device Revision or Replacement	A25	No Apparent Adverse Event	03/06/2019	
9019	2019	All other reportable incidents	47270	Cardiac resynchronization therapy implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	16/05/2019	
9020	2020	All other reportable incidents	36241	Pacemaker/defibrillator lead	CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A072201	High Impedance		
							F23	Unexpected Medical Intervention	A072201	High Impedance		
9021	2020	All other reportable incidents	37265	Dual-chamber implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A070909	Oversensing	15/09/2022	15/09/2022
									A072201	High Impedance	15/09/2022	15/09/2022
							F08	Hospitalization or Prolonged Hospitalization	A070909	Oversensing	15/09/2022	15/09/2022
									A072201	High Impedance	15/09/2022	15/09/2022
							F1905	Device Revision or Replacement	A070909	Oversensing	15/09/2022	15/09/2022
									A072201	High Impedance	15/09/2022	15/09/2022
							F23	Unexpected Medical Intervention	A070909	Oversensing	15/09/2022	15/09/2022
									A072201	High Impedance	15/09/2022	15/09/2022
9022	2020	All other reportable incidents	37265	Dual-chamber implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A070909	Oversensing		
									A0712	Pacing Problem		
									A072201	High Impedance		
							F26	No Health Consequences or Impact	A070909	Oversensing		
									A0712	Pacing Problem		
									A072201	High Impedance		
9023	2019	All other reportable incidents	35852	Single-chamber implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	23/05/2019	
9024	2019	All other reportable incidents	17882	Physiologic-monitoring defibrillation system	CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070802	Intermittent Loss of Power		
9025	2020	Death	47265	Dual-chamber implantable pacemaker, rate-responsive	CT141	Electrodes/Leads and associated devices	E2401	Insufficient Information	A070504	Premature Discharge of Battery		
							F02	Death	A070504	Premature Discharge of Battery		
							F21	Unexpected Deterioration	A070504	Premature Discharge of Battery		
					CT145	Cardiovascular devices	E2401	Insufficient Information	A070504	Premature Discharge of Battery		
							F02	Death	A070504	Premature Discharge of Battery		
							F21	Unexpected Deterioration	A070504	Premature Discharge of Battery		
					CT1684	Cardiac pacemakers and associated devices	E2401	Insufficient Information	A070504	Premature Discharge of Battery		
							F02	Death	A070504	Premature Discharge of Battery		
							F21	Unexpected Deterioration	A070504	Premature Discharge of Battery		
					CT1685	Cardiac pacemakers	E2401	Insufficient Information	A070504	Premature Discharge of Battery		
							F02	Death	A070504	Premature Discharge of Battery		
							F21	Unexpected Deterioration	A070504	Premature Discharge of Battery		
					CT233	Surgical	E2401	Insufficient Information	A070504	Premature Discharge of Battery		
							F02	Death	A070504	Premature Discharge of Battery		
							F21	Unexpected Deterioration	A070504	Premature Discharge of Battery		
					CT2406	Implantable	E2401	Insufficient Information	A070504	Premature Discharge of Battery		
							F02	Death	A070504	Premature Discharge of Battery		
							F21	Unexpected Deterioration	A070504	Premature Discharge of Battery		
					CT2407	Active implantable	E2401	Insufficient Information	A070504	Premature Discharge of Battery		
							F02	Death	A070504	Premature Discharge of Battery		
							F21	Unexpected Deterioration	A070504	Premature Discharge of Battery		
					CT257	Implantable cardiac pacemakers	E2401	Insufficient Information	A070504	Premature Discharge of Battery		
							F02	Death	A070504	Premature Discharge of Battery		

							F21	Unexpected Deterioration	A070504	Premature Discharge of Battery		
					CT2727	Active implant devices/ancillaries	E2401	Insufficient Information	A070504	Premature Discharge of Battery		
							F02	Death	A070504	Premature Discharge of Battery		
							F21	Unexpected Deterioration	A070504	Premature Discharge of Battery		
					CT2750	Implantable cardiac electrical stimulators and associated devices	E2401	Insufficient Information	A070504	Premature Discharge of Battery		
							F02	Death	A070504	Premature Discharge of Battery		
							F21	Unexpected Deterioration	A070504	Premature Discharge of Battery		
					CT2866	Implantable cardiac electrical stimulators	E2401	Insufficient Information	A070504	Premature Discharge of Battery		
							F02	Death	A070504	Premature Discharge of Battery		
							F21	Unexpected Deterioration	A070504	Premature Discharge of Battery		
					CT2985	Active	E2401	Insufficient Information	A070504	Premature Discharge of Battery		
							F02	Death	A070504	Premature Discharge of Battery		
							F21	Unexpected Deterioration	A070504	Premature Discharge of Battery		
					CT3001	Implantable metal	E2401	Insufficient Information	A070504	Premature Discharge of Battery		
							F02	Death	A070504	Premature Discharge of Battery		
							F21	Unexpected Deterioration	A070504	Premature Discharge of Battery		
					CT3002	Implantable synthetic polymer	E2401	Insufficient Information	A070504	Premature Discharge of Battery		
							F02	Death	A070504	Premature Discharge of Battery		
							F21	Unexpected Deterioration	A070504	Premature Discharge of Battery		
					CT3006	Implantable inorganic material	E2401	Insufficient Information	A070504	Premature Discharge of Battery		
							F02	Death	A070504	Premature Discharge of Battery		
							F21	Unexpected Deterioration	A070504	Premature Discharge of Battery		
					CT314	Electrically powered	E2401	Insufficient Information	A070504	Premature Discharge of Battery		
							F02	Death	A070504	Premature Discharge of Battery		
							F21	Unexpected Deterioration	A070504	Premature Discharge of Battery		
					CT321	Long-term surgical invasive	E2401	Insufficient Information	A070504	Premature Discharge of Battery		
							F02	Death	A070504	Premature Discharge of Battery		
							F21	Unexpected Deterioration	A070504	Premature Discharge of Battery		
					CT449	Generators and associated devices	E2401	Insufficient Information	A070504	Premature Discharge of Battery		
							F02	Death	A070504	Premature Discharge of Battery		
							F21	Unexpected Deterioration	A070504	Premature Discharge of Battery		
					CT981	Single use	E2401	Insufficient Information	A070504	Premature Discharge of Battery		
							F02	Death	A070504	Premature Discharge of Battery		
							F21	Unexpected Deterioration	A070504	Premature Discharge of Battery		
					CT983	Surgical invasive	E2401	Insufficient Information	A070504	Premature Discharge of Battery		
							F02	Death	A070504	Premature Discharge of Battery		
							F21	Unexpected Deterioration	A070504	Premature Discharge of Battery		
9026	2020	All other reportable incidents	35223	Endocardial/interventricular septal pacing lead	CT1684	Cardiac pacemakers and associated devices	E210401	Shock from Patient Lead(s)	A070901	Decreased Sensitivity	24/03/2020	24/03/2020
									A070909	Oversensing	24/03/2020	24/03/2020
									A071302	Inappropriate/Inadequate Shock/Stimulation	24/03/2020	24/03/2020
							F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A070901	Decreased Sensitivity	24/03/2020	24/03/2020
									A070909	Oversensing	24/03/2020	24/03/2020
									A071302	Inappropriate/Inadequate Shock/Stimulation	24/03/2020	24/03/2020
							F1203	Life Threatening Illness or Injury	A070901	Decreased Sensitivity	24/03/2020	24/03/2020
									A070909	Oversensing	24/03/2020	24/03/2020
									A071302	Inappropriate/Inadequate Shock/Stimulation	24/03/2020	24/03/2020
							F1905	Device Revision or Replacement	A070901	Decreased Sensitivity	24/03/2020	24/03/2020
									A070909	Oversensing	24/03/2020	24/03/2020
									A071302	Inappropriate/Inadequate Shock/Stimulation	24/03/2020	24/03/2020
9027	2020	All other reportable incidents	47265	Dual-chamber implantable pacemaker, rate-responsive	CT141	Electrodes/Leads and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem	07/04/2020	07/04/2020
							F1901	Additional Surgery	A0705	Battery Problem	07/04/2020	07/04/2020
					CT145	Cardiovascular devices	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem	07/04/2020	07/04/2020
							F1901	Additional Surgery	A0705	Battery Problem	07/04/2020	07/04/2020
					CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem	07/04/2020	07/04/2020
							F1901	Additional Surgery	A0705	Battery Problem	07/04/2020	07/04/2020
					CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem	07/04/2020	07/04/2020
							F1901	Additional Surgery	A0705	Battery Problem	07/04/2020	07/04/2020
					CT233	Surgical	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem	07/04/2020	07/04/2020
							F1901	Additional Surgery	A0705	Battery Problem	07/04/2020	07/04/2020
					CT2406	Implantable	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem	07/04/2020	07/04/2020
							F1901	Additional Surgery	A0705	Battery Problem	07/04/2020	07/04/2020
					CT2407	Active implantable	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem	07/04/2020	07/04/2020
							F1901	Additional Surgery	A0705	Battery Problem	07/04/2020	07/04/2020
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem	07/04/2020	07/04/2020
							F1901	Additional Surgery	A0705	Battery Problem	07/04/2020	07/04/2020
					CT2727	Active implant devices/ancillaries	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem	07/04/2020	07/04/2020
							F1901	Additional Surgery	A0705	Battery Problem	07/04/2020	07/04/2020
					CT2750	Implantable cardiac electrical stimulators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem	07/04/2020	07/04/2020
							F1901	Additional Surgery	A0705	Battery Problem	07/04/2020	07/04/2020
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem	07/04/2020	07/04/2020
							F1901	Additional Surgery	A0705	Battery Problem	07/04/2020	07/04/2020
					CT2985	Active	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem	07/04/2020	07/04/2020
							F1901	Additional Surgery	A0705	Battery Problem	07/04/2020	07/04/2020
					CT3001	Implantable metal	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem	07/04/2020	07/04/2020
							F1901	Additional Surgery	A0705	Battery Problem	07/04/2020	07/04/2020
					CT3002	Implantable synthetic polymer	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem	07/04/2020	07/04/2020
							F1901	Additional Surgery	A0705	Battery Problem	07/04/2020	07/04/2020
					CT3006	Implantable inorganic material	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem	07/04/2020	07/04/2020
							F1901	Additional Surgery	A0705	Battery Problem	07/04/2020	07/04/2020
					CT314	Electrically powered	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem	07/04/2020	07/04/2020
							F1901	Additional Surgery	A0705	Battery Problem	07/04/2020	07/04/2020
					CT321	Long-term surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem	07/04/2020	07/04/2020
							F1901	Additional Surgery	A0705	Battery Problem	07/04/2020	07/04/2020
					CT449	Generators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem	07/04/2020	07/04/2020
							F1901	Additional Surgery	A0705	Battery Problem	07/04/2020	07/04/2020
					CT981	Single use	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem	07/04/2020	07/04/2020
							F1901	Additional Surgery	A0705	Battery Problem	07/04/2020	07/04/2020
					CT983	Surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem	07/04/2020	07/04/2020

							F1901	Additional Surgery	A0705	Battery Problem	07/04/2020	07/04/2020
9028	2020	All other reportable incidents	47265	Dual-chamber implantable pacemaker, rate-responsive	CT141	Electrodes/Leads and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem	26/03/2020	26/03/2020
									A070504	Premature Discharge of Battery	26/03/2020	26/03/2020
							F08	Hospitalization or Prolonged Hospitalization	A0705	Battery Problem	26/03/2020	26/03/2020
									A070504	Premature Discharge of Battery	26/03/2020	26/03/2020
							F12	Serious Injury/ Illness/ Impairment	A0705	Battery Problem	26/03/2020	26/03/2020
									A070504	Premature Discharge of Battery	26/03/2020	26/03/2020
							F19	Surgical Intervention	A0705	Battery Problem	26/03/2020	26/03/2020
									A070504	Premature Discharge of Battery	26/03/2020	26/03/2020
					CT145	Cardiovascular devices	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem	26/03/2020	26/03/2020
									A070504	Premature Discharge of Battery	26/03/2020	26/03/2020
							F08	Hospitalization or Prolonged Hospitalization	A0705	Battery Problem	26/03/2020	26/03/2020
									A070504	Premature Discharge of Battery	26/03/2020	26/03/2020
							F12	Serious Injury/ Illness/ Impairment	A0705	Battery Problem	26/03/2020	26/03/2020
									A070504	Premature Discharge of Battery	26/03/2020	26/03/2020
							F19	Surgical Intervention	A0705	Battery Problem	26/03/2020	26/03/2020
									A070504	Premature Discharge of Battery	26/03/2020	26/03/2020
					CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem	26/03/2020	26/03/2020
									A070504	Premature Discharge of Battery	26/03/2020	26/03/2020
							F08	Hospitalization or Prolonged Hospitalization	A0705	Battery Problem	26/03/2020	26/03/2020
									A070504	Premature Discharge of Battery	26/03/2020	26/03/2020
							F12	Serious Injury/ Illness/ Impairment	A0705	Battery Problem	26/03/2020	26/03/2020
									A070504	Premature Discharge of Battery	26/03/2020	26/03/2020
							F19	Surgical Intervention	A0705	Battery Problem	26/03/2020	26/03/2020
									A070504	Premature Discharge of Battery	26/03/2020	26/03/2020
					CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem	26/03/2020	26/03/2020
									A070504	Premature Discharge of Battery	26/03/2020	26/03/2020
							F08	Hospitalization or Prolonged Hospitalization	A0705	Battery Problem	26/03/2020	26/03/2020
									A070504	Premature Discharge of Battery	26/03/2020	26/03/2020
							F12	Serious Injury/ Illness/ Impairment	A0705	Battery Problem	26/03/2020	26/03/2020
									A070504	Premature Discharge of Battery	26/03/2020	26/03/2020
							F19	Surgical Intervention	A0705	Battery Problem	26/03/2020	26/03/2020
									A070504	Premature Discharge of Battery	26/03/2020	26/03/2020
					CT233	Surgical	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem	26/03/2020	26/03/2020
									A070504	Premature Discharge of Battery	26/03/2020	26/03/2020
							F08	Hospitalization or Prolonged Hospitalization	A0705	Battery Problem	26/03/2020	26/03/2020
									A070504	Premature Discharge of Battery	26/03/2020	26/03/2020
							F12	Serious Injury/ Illness/ Impairment	A0705	Battery Problem	26/03/2020	26/03/2020
									A070504	Premature Discharge of Battery	26/03/2020	26/03/2020
							F19	Surgical Intervention	A0705	Battery Problem	26/03/2020	26/03/2020
									A070504	Premature Discharge of Battery	26/03/2020	26/03/2020
					CT2406	Implantable	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem	26/03/2020	26/03/2020
									A070504	Premature Discharge of Battery	26/03/2020	26/03/2020
							F08	Hospitalization or Prolonged Hospitalization	A0705	Battery Problem	26/03/2020	26/03/2020
									A070504	Premature Discharge of Battery	26/03/2020	26/03/2020
							F12	Serious Injury/ Illness/ Impairment	A0705	Battery Problem	26/03/2020	26/03/2020
									A070504	Premature Discharge of Battery	26/03/2020	26/03/2020
							F19	Surgical Intervention	A0705	Battery Problem	26/03/2020	26/03/2020
									A070504	Premature Discharge of Battery	26/03/2020	26/03/2020
					CT2407	Active implantable	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem	26/03/2020	26/03/2020
									A070504	Premature Discharge of Battery	26/03/2020	26/03/2020
							F08	Hospitalization or Prolonged Hospitalization	A0705	Battery Problem	26/03/2020	26/03/2020
									A070504	Premature Discharge of Battery	26/03/2020	26/03/2020
							F12	Serious Injury/ Illness/ Impairment	A0705	Battery Problem	26/03/2020	26/03/2020
									A070504	Premature Discharge of Battery	26/03/2020	26/03/2020
							F19	Surgical Intervention	A0705	Battery Problem	26/03/2020	26/03/2020
									A070504	Premature Discharge of Battery	26/03/2020	26/03/2020
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem	26/03/2020	26/03/2020
									A070504	Premature Discharge of Battery	26/03/2020	26/03/2020
							F08	Hospitalization or Prolonged Hospitalization	A0705	Battery Problem	26/03/2020	26/03/2020
									A070504	Premature Discharge of Battery	26/03/2020	26/03/2020
							F12	Serious Injury/ Illness/ Impairment	A0705	Battery Problem	26/03/2020	26/03/2020
									A070504	Premature Discharge of Battery	26/03/2020	26/03/2020
							F19	Surgical Intervention	A0705	Battery Problem	26/03/2020	26/03/2020
									A070504	Premature Discharge of Battery	26/03/2020	26/03/2020
					CT2727	Active implant devices/ancillaries	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem	26/03/2020	26/03/2020
									A070504	Premature Discharge of Battery	26/03/2020	26/03/2020
							F08	Hospitalization or Prolonged Hospitalization	A0705	Battery Problem	26/03/2020	26/03/2020
									A070504	Premature Discharge of Battery	26/03/2020	26/03/2020
							F12	Serious Injury/ Illness/ Impairment	A0705	Battery Problem	26/03/2020	26/03/2020
									A070504	Premature Discharge of Battery	26/03/2020	26/03/2020
							F19	Surgical Intervention	A0705	Battery Problem	26/03/2020	26/03/2020
									A070504	Premature Discharge of Battery	26/03/2020	26/03/2020
					CT2750	Implantable cardiac electrical stimulators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem	26/03/2020	26/03/2020
									A070504	Premature Discharge of Battery	26/03/2020	26/03/2020
							F08	Hospitalization or Prolonged Hospitalization	A0705	Battery Problem	26/03/2020	26/03/2020
									A070504	Premature Discharge of Battery	26/03/2020	26/03/2020
							F12	Serious Injury/ Illness/ Impairment	A0705	Battery Problem	26/03/2020	26/03/2020
									A070504	Premature Discharge of Battery	26/03/2020	26/03/2020
							F19	Surgical Intervention	A0705	Battery Problem	26/03/2020	26/03/2020
									A070504	Premature Discharge of Battery	26/03/2020	26/03/2020
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem	26/03/2020	26/03/2020
									A070504	Premature Discharge of Battery	26/03/2020	26/03/2020
							F08	Hospitalization or Prolonged Hospitalization	A0705	Battery Problem	26/03/2020	26/03/2020
									A070504	Premature Discharge of Battery	26/03/2020	26/03/2020
							F12	Serious Injury/ Illness/ Impairment	A0705	Battery Problem	26/03/2020	26/03/2020
									A070504	Premature Discharge of Battery	26/03/2020	26/03/2020
							F19	Surgical Intervention	A0705	Battery Problem	26/03/2020	26/03/2020

									A070504	Premature Discharge of Battery	26/03/2020	26/03/2020
					CT2985	Active	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem	26/03/2020	26/03/2020
									A070504	Premature Discharge of Battery	26/03/2020	26/03/2020
							F08	Hospitalization or Prolonged Hospitalization	A0705	Battery Problem	26/03/2020	26/03/2020
									A070504	Premature Discharge of Battery	26/03/2020	26/03/2020
							F12	Serious Injury/ Illness/ Impairment	A0705	Battery Problem	26/03/2020	26/03/2020
									A070504	Premature Discharge of Battery	26/03/2020	26/03/2020
							F19	Surgical Intervention	A0705	Battery Problem	26/03/2020	26/03/2020
									A070504	Premature Discharge of Battery	26/03/2020	26/03/2020
					CT3001	Implantable metal	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem	26/03/2020	26/03/2020
									A070504	Premature Discharge of Battery	26/03/2020	26/03/2020
							F08	Hospitalization or Prolonged Hospitalization	A0705	Battery Problem	26/03/2020	26/03/2020
									A070504	Premature Discharge of Battery	26/03/2020	26/03/2020
							F12	Serious Injury/ Illness/ Impairment	A0705	Battery Problem	26/03/2020	26/03/2020
									A070504	Premature Discharge of Battery	26/03/2020	26/03/2020
							F19	Surgical Intervention	A0705	Battery Problem	26/03/2020	26/03/2020
									A070504	Premature Discharge of Battery	26/03/2020	26/03/2020
					CT3002	Implantable synthetic polymer	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem	26/03/2020	26/03/2020
									A070504	Premature Discharge of Battery	26/03/2020	26/03/2020
							F08	Hospitalization or Prolonged Hospitalization	A0705	Battery Problem	26/03/2020	26/03/2020
									A070504	Premature Discharge of Battery	26/03/2020	26/03/2020
							F12	Serious Injury/ Illness/ Impairment	A0705	Battery Problem	26/03/2020	26/03/2020
									A070504	Premature Discharge of Battery	26/03/2020	26/03/2020
							F19	Surgical Intervention	A0705	Battery Problem	26/03/2020	26/03/2020
									A070504	Premature Discharge of Battery	26/03/2020	26/03/2020
					CT3006	Implantable inorganic material	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem	26/03/2020	26/03/2020
									A070504	Premature Discharge of Battery	26/03/2020	26/03/2020
							F08	Hospitalization or Prolonged Hospitalization	A0705	Battery Problem	26/03/2020	26/03/2020
									A070504	Premature Discharge of Battery	26/03/2020	26/03/2020
							F12	Serious Injury/ Illness/ Impairment	A0705	Battery Problem	26/03/2020	26/03/2020
									A070504	Premature Discharge of Battery	26/03/2020	26/03/2020
							F19	Surgical Intervention	A0705	Battery Problem	26/03/2020	26/03/2020
									A070504	Premature Discharge of Battery	26/03/2020	26/03/2020
					CT314	Electrically powered	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem	26/03/2020	26/03/2020
									A070504	Premature Discharge of Battery	26/03/2020	26/03/2020
							F08	Hospitalization or Prolonged Hospitalization	A0705	Battery Problem	26/03/2020	26/03/2020
									A070504	Premature Discharge of Battery	26/03/2020	26/03/2020
							F12	Serious Injury/ Illness/ Impairment	A0705	Battery Problem	26/03/2020	26/03/2020
									A070504	Premature Discharge of Battery	26/03/2020	26/03/2020
							F19	Surgical Intervention	A0705	Battery Problem	26/03/2020	26/03/2020
									A070504	Premature Discharge of Battery	26/03/2020	26/03/2020
					CT321	Long-term surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem	26/03/2020	26/03/2020
									A070504	Premature Discharge of Battery	26/03/2020	26/03/2020
							F08	Hospitalization or Prolonged Hospitalization	A0705	Battery Problem	26/03/2020	26/03/2020
									A070504	Premature Discharge of Battery	26/03/2020	26/03/2020
							F12	Serious Injury/ Illness/ Impairment	A0705	Battery Problem	26/03/2020	26/03/2020
									A070504	Premature Discharge of Battery	26/03/2020	26/03/2020
							F19	Surgical Intervention	A0705	Battery Problem	26/03/2020	26/03/2020
									A070504	Premature Discharge of Battery	26/03/2020	26/03/2020
					CT449	Generators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem	26/03/2020	26/03/2020
									A070504	Premature Discharge of Battery	26/03/2020	26/03/2020
							F08	Hospitalization or Prolonged Hospitalization	A0705	Battery Problem	26/03/2020	26/03/2020
									A070504	Premature Discharge of Battery	26/03/2020	26/03/2020
							F12	Serious Injury/ Illness/ Impairment	A0705	Battery Problem	26/03/2020	26/03/2020
									A070504	Premature Discharge of Battery	26/03/2020	26/03/2020
							F19	Surgical Intervention	A0705	Battery Problem	26/03/2020	26/03/2020
									A070504	Premature Discharge of Battery	26/03/2020	26/03/2020
					CT981	Single use	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem	26/03/2020	26/03/2020
									A070504	Premature Discharge of Battery	26/03/2020	26/03/2020
							F08	Hospitalization or Prolonged Hospitalization	A0705	Battery Problem	26/03/2020	26/03/2020
									A070504	Premature Discharge of Battery	26/03/2020	26/03/2020
							F12	Serious Injury/ Illness/ Impairment	A0705	Battery Problem	26/03/2020	26/03/2020
									A070504	Premature Discharge of Battery	26/03/2020	26/03/2020
							F19	Surgical Intervention	A0705	Battery Problem	26/03/2020	26/03/2020
									A070504	Premature Discharge of Battery	26/03/2020	26/03/2020
					CT983	Surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem	26/03/2020	26/03/2020
									A070504	Premature Discharge of Battery	26/03/2020	26/03/2020
							F08	Hospitalization or Prolonged Hospitalization	A0705	Battery Problem	26/03/2020	26/03/2020
									A070504	Premature Discharge of Battery	26/03/2020	26/03/2020
							F12	Serious Injury/ Illness/ Impairment	A0705	Battery Problem	26/03/2020	26/03/2020
									A070504	Premature Discharge of Battery	26/03/2020	26/03/2020
							F19	Surgical Intervention	A0705	Battery Problem	26/03/2020	26/03/2020
									A070504	Premature Discharge of Battery	26/03/2020	26/03/2020
9029	2020	All other reportable incidents	17882	Physiologic-monitoring defibrillation system	CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A071301	Failure to Deliver Shock/Stimulation		
							F27	Problem Identified During Non-Clinical Procedure	A071301	Failure to Deliver Shock/Stimulation		
9030	2020	All other reportable incidents	17882	Physiologic-monitoring defibrillation system	CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070912	Incorrect Interpretation of Signal		
							F26	No Health Consequences or Impact	A070912	Incorrect Interpretation of Signal		
9031	2019	All other reportable incidents	36241	Pacemaker/defibrillator lead	CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A0510	Retraction Problem	02/05/2019	
9032	2019	All other reportable incidents	36241	Pacemaker/defibrillator lead	CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A040101	Fracture	31/05/2019	31/05/2019
									A070909	Oversensing	31/05/2019	31/05/2019
									A071302	Inappropriate/Inadequate Shock/Stimulation	31/05/2019	31/05/2019
							F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A040101	Fracture	31/05/2019	31/05/2019
									A070909	Oversensing	31/05/2019	31/05/2019
									A071302	Inappropriate/Inadequate Shock/Stimulation	31/05/2019	31/05/2019
							F1903	Device Explantation	A040101	Fracture	31/05/2019	31/05/2019
									A070909	Oversensing	31/05/2019	31/05/2019
									A071302	Inappropriate/Inadequate Shock/Stimulation	31/05/2019	31/05/2019
							F1905	Device Revision or Replacement	A040101	Fracture	31/05/2019	31/05/2019

									A070909	Oversensing	31/05/2019	31/05/2019
									A071302	Inappropriate/Inadequate Shock/Stimulation	31/05/2019	31/05/2019
9033	2020	All other reportable incidents	35852	Single-chamber implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A070909	Oversensing		
9034	2020	All other reportable incidents	35852	Single-chamber implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A071302	Inappropriate/Inadequate Shock/Stimulation		
9035	2020	All other reportable incidents	47267	Single-chamber implantable pacemaker, rate-responsive	CT141	Electrodes/Leads and associated devices	E0601	Arrhythmia	A070504	Premature Discharge of Battery	02/03/2020	02/03/2020
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	02/03/2020	02/03/2020
					CT145	Cardiovascular devices	E0601	Arrhythmia	A070504	Premature Discharge of Battery	02/03/2020	02/03/2020
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	02/03/2020	02/03/2020
					CT1684	Cardiac pacemakers and associated devices	E0601	Arrhythmia	A070504	Premature Discharge of Battery	02/03/2020	02/03/2020
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	02/03/2020	02/03/2020
					CT1685	Cardiac pacemakers	E0601	Arrhythmia	A070504	Premature Discharge of Battery	02/03/2020	02/03/2020
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	02/03/2020	02/03/2020
					CT233	Surgical	E0601	Arrhythmia	A070504	Premature Discharge of Battery	02/03/2020	02/03/2020
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	02/03/2020	02/03/2020
					CT2406	Implantable	E0601	Arrhythmia	A070504	Premature Discharge of Battery	02/03/2020	02/03/2020
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	02/03/2020	02/03/2020
					CT2407	Active implantable	E0601	Arrhythmia	A070504	Premature Discharge of Battery	02/03/2020	02/03/2020
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	02/03/2020	02/03/2020
					CT257	Implantable cardiac pacemakers	E0601	Arrhythmia	A070504	Premature Discharge of Battery	02/03/2020	02/03/2020
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	02/03/2020	02/03/2020
					CT2727	Active implant devices/ancillaries	E0601	Arrhythmia	A070504	Premature Discharge of Battery	02/03/2020	02/03/2020
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	02/03/2020	02/03/2020
					CT2750	Implantable cardiac electrical stimulators and associated devices	E0601	Arrhythmia	A070504	Premature Discharge of Battery	02/03/2020	02/03/2020
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	02/03/2020	02/03/2020
					CT2866	Implantable cardiac electrical stimulators	E0601	Arrhythmia	A070504	Premature Discharge of Battery	02/03/2020	02/03/2020
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	02/03/2020	02/03/2020
					CT2985	Active	E0601	Arrhythmia	A070504	Premature Discharge of Battery	02/03/2020	02/03/2020
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	02/03/2020	02/03/2020
					CT3001	Implantable metal	E0601	Arrhythmia	A070504	Premature Discharge of Battery	02/03/2020	02/03/2020
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	02/03/2020	02/03/2020
					CT3002	Implantable synthetic polymer	E0601	Arrhythmia	A070504	Premature Discharge of Battery	02/03/2020	02/03/2020
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	02/03/2020	02/03/2020
					CT3006	Implantable inorganic material	E0601	Arrhythmia	A070504	Premature Discharge of Battery	02/03/2020	02/03/2020
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	02/03/2020	02/03/2020
					CT314	Electrically powered	E0601	Arrhythmia	A070504	Premature Discharge of Battery	02/03/2020	02/03/2020
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	02/03/2020	02/03/2020
					CT321	Long-term surgical invasive	E0601	Arrhythmia	A070504	Premature Discharge of Battery	02/03/2020	02/03/2020
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	02/03/2020	02/03/2020
					CT449	Generators and associated devices	E0601	Arrhythmia	A070504	Premature Discharge of Battery	02/03/2020	02/03/2020
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	02/03/2020	02/03/2020
					CT981	Single use	E0601	Arrhythmia	A070504	Premature Discharge of Battery	02/03/2020	02/03/2020
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	02/03/2020	02/03/2020
					CT983	Surgical invasive	E0601	Arrhythmia	A070504	Premature Discharge of Battery	02/03/2020	02/03/2020
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	02/03/2020	02/03/2020
9036	2020	All other reportable incidents	36241	Pacemaker/defibrillator lead	CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070909	Oversensing		
									A0712	Pacing Problem		
9037	2020	All other reportable incidents	35852	Single-chamber implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A070909	Oversensing		
							F26	No Health Consequences or Impact	A070909	Oversensing		
9038	2020	All other reportable incidents	47265	Dual-chamber implantable pacemaker, rate-responsive	CT141	Electrodes/Leads and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
					CT145	Cardiovascular devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
					CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
					CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
					CT233	Surgical	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
					CT2406	Implantable	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
					CT2407	Active implantable	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
					CT2727	Active implant devices/ancillaries	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
					CT2750	Implantable cardiac electrical stimulators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
					CT2985	Active	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
					CT3001	Implantable metal	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
					CT3002	Implantable synthetic polymer	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
					CT3006	Implantable inorganic material	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
					CT314	Electrically powered	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
					CT321	Long-term surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
					CT449	Generators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
					CT981	Single use	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
					CT983	Surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
9039	2020	All other reportable incidents	47267	Single-chamber implantable pacemaker, rate-responsive	CT141	Electrodes/Leads and associated devices	F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery		
					CT145	Cardiovascular devices	F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery		
					CT1684	Cardiac pacemakers and associated devices	F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery		
					CT1685	Cardiac pacemakers	F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery		
					CT233	Surgical	F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery		
					CT2406	Implantable	F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery		
					CT2407	Active implantable	F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery		
					CT257	Implantable cardiac pacemakers	F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery		
					CT2727	Active implant devices/ancillaries	F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery		
					CT2750	Implantable cardiac electrical stimulators and associated devices	F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery		
					CT2866	Implantable cardiac electrical stimulators	F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery		
					CT2985	Active	F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery		
					CT3001	Implantable metal	F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery		
					CT3002	Implantable synthetic polymer	F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery		
					CT3006	Implantable inorganic material	F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery		
					CT314	Electrically powered	F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery		
					CT321	Long-term surgical invasive	F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery		
					CT449	Generators and associated devices	F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery		

					CT981	Single use	F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery		
					CT983	Surgical invasive	F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery		
9040	2020	All other reportable incidents	47265	Dual-chamber implantable pacemaker, rate-responsive	CT141	Electrodes/Leads and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT145	Cardiovascular devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT233	Surgical	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT2406	Implantable	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT2407	Active implantable	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT2727	Active implant devices/ancillaries	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT2750	Implantable cardiac electrical stimulators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT2985	Active	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT3001	Implantable metal	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT3002	Implantable synthetic polymer	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT3006	Implantable inorganic material	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT314	Electrically powered	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT321	Long-term surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT449	Generators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT981	Single use	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT983	Surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
9041	2020	All other reportable incidents	47263	Cardiac resynchronization therapy implantable pacemaker	CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A072201	High Impedance		
							F26	No Health Consequences or Impact	A072201	High Impedance		
					CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A072201	High Impedance		
							F26	No Health Consequences or Impact	A072201	High Impedance		
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A072201	High Impedance		
							F26	No Health Consequences or Impact	A072201	High Impedance		
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A072201	High Impedance		
							F26	No Health Consequences or Impact	A072201	High Impedance		
9042	2020	All other reportable incidents	47265	Dual-chamber implantable pacemaker, rate-responsive	CT141	Electrodes/Leads and associated devices	E2312	Fatigue	A070504	Premature Discharge of Battery		
							F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A070504	Premature Discharge of Battery		
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery		
					CT145	Cardiovascular devices	E2312	Fatigue	A070504	Premature Discharge of Battery		
							F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A070504	Premature Discharge of Battery		
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery		
					CT1684	Cardiac pacemakers and associated devices	E2312	Fatigue	A070504	Premature Discharge of Battery		
							F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A070504	Premature Discharge of Battery		
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery		
					CT1685	Cardiac pacemakers	E2312	Fatigue	A070504	Premature Discharge of Battery		
							F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A070504	Premature Discharge of Battery		
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery		
					CT233	Surgical	E2312	Fatigue	A070504	Premature Discharge of Battery		
							F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A070504	Premature Discharge of Battery		
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery		
					CT2406	Implantable	E2312	Fatigue	A070504	Premature Discharge of Battery		
							F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A070504	Premature Discharge of Battery		
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery		
					CT2407	Active implantable	E2312	Fatigue	A070504	Premature Discharge of Battery		
							F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A070504	Premature Discharge of Battery		
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery		
					CT257	Implantable cardiac pacemakers	E2312	Fatigue	A070504	Premature Discharge of Battery		
							F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A070504	Premature Discharge of Battery		
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery		
					CT2727	Active implant devices/ancillaries	E2312	Fatigue	A070504	Premature Discharge of Battery		
							F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A070504	Premature Discharge of Battery		
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery		
					CT2750	Implantable cardiac electrical stimulators and associated devices	E2312	Fatigue	A070504	Premature Discharge of Battery		
							F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A070504	Premature Discharge of Battery		
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery		
					CT2866	Implantable cardiac electrical stimulators	E2312	Fatigue	A070504	Premature Discharge of Battery		
							F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A070504	Premature Discharge of Battery		
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery		
					CT2985	Active	E2312	Fatigue	A070504	Premature Discharge of Battery		
							F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A070504	Premature Discharge of Battery		
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery		
					CT3001	Implantable metal	E2312	Fatigue	A070504	Premature Discharge of Battery		
							F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A070504	Premature Discharge of Battery		

						F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery		
					CT3002	Implantable synthetic polymer	E2312	Fatigue	A070504	Premature Discharge of Battery		
							F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A070504	Premature Discharge of Battery		
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery		
					CT3006	Implantable inorganic material	E2312	Fatigue	A070504	Premature Discharge of Battery		
							F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A070504	Premature Discharge of Battery		
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery		
					CT314	Electrically powered	E2312	Fatigue	A070504	Premature Discharge of Battery		
							F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A070504	Premature Discharge of Battery		
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery		
					CT321	Long-term surgical invasive	E2312	Fatigue	A070504	Premature Discharge of Battery		
							F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A070504	Premature Discharge of Battery		
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery		
					CT449	Generators and associated devices	E2312	Fatigue	A070504	Premature Discharge of Battery		
							F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A070504	Premature Discharge of Battery		
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery		
					CT981	Single use	E2312	Fatigue	A070504	Premature Discharge of Battery		
							F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A070504	Premature Discharge of Battery		
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery		
					CT983	Surgical invasive	E2312	Fatigue	A070504	Premature Discharge of Battery		
							F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A070504	Premature Discharge of Battery		
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery		
9043	2020	All other reportable incidents	47265	Dual-chamber implantable pacemaker, rate-responsive	CT141	Electrodes/Leads and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F23	Unexpected Medical Intervention	A070504	Premature Discharge of Battery		
					CT145	Cardiovascular devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F23	Unexpected Medical Intervention	A070504	Premature Discharge of Battery		
					CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F23	Unexpected Medical Intervention	A070504	Premature Discharge of Battery		
					CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F23	Unexpected Medical Intervention	A070504	Premature Discharge of Battery		
					CT233	Surgical	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F23	Unexpected Medical Intervention	A070504	Premature Discharge of Battery		
					CT2406	Implantable	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F23	Unexpected Medical Intervention	A070504	Premature Discharge of Battery		
					CT2407	Active implantable	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F23	Unexpected Medical Intervention	A070504	Premature Discharge of Battery		
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F23	Unexpected Medical Intervention	A070504	Premature Discharge of Battery		
					CT2727	Active implant devices/ancillaries	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F23	Unexpected Medical Intervention	A070504	Premature Discharge of Battery		
					CT2750	Implantable cardiac electrical stimulators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F23	Unexpected Medical Intervention	A070504	Premature Discharge of Battery		
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F23	Unexpected Medical Intervention	A070504	Premature Discharge of Battery		
					CT2985	Active	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F23	Unexpected Medical Intervention	A070504	Premature Discharge of Battery		
					CT3001	Implantable metal	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F23	Unexpected Medical Intervention	A070504	Premature Discharge of Battery		
					CT3002	Implantable synthetic polymer	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F23	Unexpected Medical Intervention	A070504	Premature Discharge of Battery		
					CT3006	Implantable inorganic material	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F23	Unexpected Medical Intervention	A070504	Premature Discharge of Battery		
					CT314	Electrically powered	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F23	Unexpected Medical Intervention	A070504	Premature Discharge of Battery		
					CT321	Long-term surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F23	Unexpected Medical Intervention	A070504	Premature Discharge of Battery		
					CT449	Generators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F23	Unexpected Medical Intervention	A070504	Premature Discharge of Battery		
					CT981	Single use	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F23	Unexpected Medical Intervention	A070504	Premature Discharge of Battery		
					CT983	Surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F23	Unexpected Medical Intervention	A070504	Premature Discharge of Battery		
9044	2019	All other reportable incidents	17882	Physiologic-monitoring defibrillation system	CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070803	Failure to Power Up		
							F27	Problem Identified During Non-Clinical Procedure	A070803	Failure to Power Up		
9045	2020	All other reportable incidents	36241	Pacemaker/defibrillator lead	CT1684	Cardiac pacemakers and associated devices	E2007	Fall	A070909	Oversensing		
									A0712	Pacing Problem		
									A072201	High Impedance		
							E2403	No Clinical Signs, Symptoms or Conditions	A070909	Oversensing		
									A0712	Pacing Problem		
									A072201	High Impedance		
							F1905	Device Revision or Replacement	A070909	Oversensing		
									A0712	Pacing Problem		
									A072201	High Impedance		
							F24	Insufficient Health Impact Information	A070909	Oversensing		
									A0712	Pacing Problem		
									A072201	High Impedance		
9046	2020	All other reportable incidents	35852	Single-chamber implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	30/03/2020	30/03/2020
							F1901	Additional Surgery	A070504	Premature Discharge of Battery	30/03/2020	30/03/2020
9047	2020	All other reportable incidents	35852	Single-chamber implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	01/04/2020	01/04/2020
							F1901	Additional Surgery	A070504	Premature Discharge of Battery	01/04/2020	01/04/2020
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	01/04/2020	01/04/2020
9048	2020	All other reportable incidents	17882	Physiologic-monitoring defibrillation system	CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070802	Intermittent Loss of Power		
							F26	No Health Consequences or Impact	A070802	Intermittent Loss of Power		
9049	2020	All other reportable incidents	17882	Physiologic-monitoring defibrillation system	CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A07	Electrical /Electronic Property Problem		
							F27	Problem Identified During Non-Clinical Procedure	A07	Electrical /Electronic Property Problem		
9050	2019	All other reportable incidents	47270	Cardiac resynchronization therapy implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	16/05/2019	
9051	2020	All other reportable incidents	47265	Dual-chamber implantable pacemaker, rate-	CT141	Electrodes/Leads and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		

				responsive	CT145	Cardiovascular devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
					CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
					CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
					CT233	Surgical	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
					CT2406	Implantable	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
					CT2407	Active implantable	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
					CT2727	Active implant devices/ancillaries	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
					CT2750	Implantable cardiac electrical stimulators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
					CT2985	Active	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
					CT3001	Implantable metal	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
					CT3002	Implantable synthetic polymer	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
					CT3006	Implantable inorganic material	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
					CT314	Electrically powered	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
					CT321	Long-term surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
					CT449	Generators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
					CT981	Single use	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
					CT983	Surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
9052	2019	All other reportable incidents	47265	Dual-chamber implantable pacemaker, rate-responsive	CT141	Electrodes/Leads and associated devices	F05	Delay to Treatment/ Therapy	A071206	Pacing Asynchronously	03/06/2019	
							F1905	Device Revision or Replacement	A071206	Pacing Asynchronously	03/06/2019	
					CT145	Cardiovascular devices	F05	Delay to Treatment/ Therapy	A071206	Pacing Asynchronously	03/06/2019	
							F1905	Device Revision or Replacement	A071206	Pacing Asynchronously	03/06/2019	
					CT1684	Cardiac pacemakers and associated devices	F05	Delay to Treatment/ Therapy	A071206	Pacing Asynchronously	03/06/2019	
							F1905	Device Revision or Replacement	A071206	Pacing Asynchronously	03/06/2019	
					CT1685	Cardiac pacemakers	F05	Delay to Treatment/ Therapy	A071206	Pacing Asynchronously	03/06/2019	
							F1905	Device Revision or Replacement	A071206	Pacing Asynchronously	03/06/2019	
					CT233	Surgical	F05	Delay to Treatment/ Therapy	A071206	Pacing Asynchronously	03/06/2019	
							F1905	Device Revision or Replacement	A071206	Pacing Asynchronously	03/06/2019	
					CT2406	Implantable	F05	Delay to Treatment/ Therapy	A071206	Pacing Asynchronously	03/06/2019	
							F1905	Device Revision or Replacement	A071206	Pacing Asynchronously	03/06/2019	
					CT2407	Active implantable	F05	Delay to Treatment/ Therapy	A071206	Pacing Asynchronously	03/06/2019	
							F1905	Device Revision or Replacement	A071206	Pacing Asynchronously	03/06/2019	
					CT257	Implantable cardiac pacemakers	F05	Delay to Treatment/ Therapy	A071206	Pacing Asynchronously	03/06/2019	
							F1905	Device Revision or Replacement	A071206	Pacing Asynchronously	03/06/2019	
					CT2727	Active implant devices/ancillaries	F05	Delay to Treatment/ Therapy	A071206	Pacing Asynchronously	03/06/2019	
							F1905	Device Revision or Replacement	A071206	Pacing Asynchronously	03/06/2019	
					CT2750	Implantable cardiac electrical stimulators and associated devices	F05	Delay to Treatment/ Therapy	A071206	Pacing Asynchronously	03/06/2019	
							F1905	Device Revision or Replacement	A071206	Pacing Asynchronously	03/06/2019	
					CT2866	Implantable cardiac electrical stimulators	F05	Delay to Treatment/ Therapy	A071206	Pacing Asynchronously	03/06/2019	
							F1905	Device Revision or Replacement	A071206	Pacing Asynchronously	03/06/2019	
					CT2985	Active	F05	Delay to Treatment/ Therapy	A071206	Pacing Asynchronously	03/06/2019	
							F1905	Device Revision or Replacement	A071206	Pacing Asynchronously	03/06/2019	
					CT3001	Implantable metal	F05	Delay to Treatment/ Therapy	A071206	Pacing Asynchronously	03/06/2019	
							F1905	Device Revision or Replacement	A071206	Pacing Asynchronously	03/06/2019	
					CT3002	Implantable synthetic polymer	F05	Delay to Treatment/ Therapy	A071206	Pacing Asynchronously	03/06/2019	
							F1905	Device Revision or Replacement	A071206	Pacing Asynchronously	03/06/2019	
					CT3006	Implantable inorganic material	F05	Delay to Treatment/ Therapy	A071206	Pacing Asynchronously	03/06/2019	
							F1905	Device Revision or Replacement	A071206	Pacing Asynchronously	03/06/2019	
					CT314	Electrically powered	F05	Delay to Treatment/ Therapy	A071206	Pacing Asynchronously	03/06/2019	
							F1905	Device Revision or Replacement	A071206	Pacing Asynchronously	03/06/2019	
					CT321	Long-term surgical invasive	F05	Delay to Treatment/ Therapy	A071206	Pacing Asynchronously	03/06/2019	
							F1905	Device Revision or Replacement	A071206	Pacing Asynchronously	03/06/2019	
					CT449	Generators and associated devices	F05	Delay to Treatment/ Therapy	A071206	Pacing Asynchronously	03/06/2019	
							F1905	Device Revision or Replacement	A071206	Pacing Asynchronously	03/06/2019	
					CT981	Single use	F05	Delay to Treatment/ Therapy	A071206	Pacing Asynchronously	03/06/2019	
							F1905	Device Revision or Replacement	A071206	Pacing Asynchronously	03/06/2019	
					CT983	Surgical invasive	F05	Delay to Treatment/ Therapy	A071206	Pacing Asynchronously	03/06/2019	
							F1905	Device Revision or Replacement	A071206	Pacing Asynchronously	03/06/2019	
9053	2019	All other reportable incidents	36241	Pacemaker/defibrillator lead	CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A072201	High Impedance		
9054	2020	All other reportable incidents	36241	Pacemaker/defibrillator lead	CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A05	Mechanical Problem	10/03/2020	10/03/2020
									A0709	Device Sensing Problem	10/03/2020	10/03/2020
							F26	No Health Consequences or Impact	A05	Mechanical Problem	10/03/2020	10/03/2020
									A0709	Device Sensing Problem	10/03/2020	10/03/2020
9055	2019	All other reportable incidents	47267	Single-chamber implantable pacemaker, rate-responsive	CT141	Electrodes/Leads and associated devices	F1905	Device Revision or Replacement	A0705	Battery Problem		
					CT145	Cardiovascular devices	F1905	Device Revision or Replacement	A0705	Battery Problem		
					CT1684	Cardiac pacemakers and associated devices	F1905	Device Revision or Replacement	A0705	Battery Problem		
					CT1685	Cardiac pacemakers	F1905	Device Revision or Replacement	A0705	Battery Problem		
					CT233	Surgical	F1905	Device Revision or Replacement	A0705	Battery Problem		
					CT2406	Implantable	F1905	Device Revision or Replacement	A0705	Battery Problem		
					CT2407	Active implantable	F1905	Device Revision or Replacement	A0705	Battery Problem		
					CT257	Implantable cardiac pacemakers	F1905	Device Revision or Replacement	A0705	Battery Problem		
					CT2727	Active implant devices/ancillaries	F1905	Device Revision or Replacement	A0705	Battery Problem		
					CT2750	Implantable cardiac electrical stimulators and associated devices	F1905	Device Revision or Replacement	A0705	Battery Problem		
					CT2866	Implantable cardiac electrical stimulators	F1905	Device Revision or Replacement	A0705	Battery Problem		
					CT2985	Active	F1905	Device Revision or Replacement	A0705	Battery Problem		
					CT3001	Implantable metal	F1905	Device Revision or Replacement	A0705	Battery Problem		
					CT3002	Implantable synthetic polymer	F1905	Device Revision or Replacement	A0705	Battery Problem		
					CT3006	Implantable inorganic material	F1905	Device Revision or Replacement	A0705	Battery Problem		
					CT314	Electrically powered	F1905	Device Revision or Replacement	A0705	Battery Problem		
					CT321	Long-term surgical invasive	F1905	Device Revision or Replacement	A0705	Battery Problem		
					CT449	Generators and associated devices	F1905	Device Revision or Replacement	A0705	Battery Problem		
					CT981	Single use	F1905	Device Revision or Replacement	A0705	Battery Problem		
					CT983	Surgical invasive	F1905	Device Revision or Replacement	A0705	Battery Problem		
9056	2019	All other reportable incidents	37265	Dual-chamber implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		

9057	2019	All other reportable incidents	47270	Cardiac resynchronization therapy implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	24/05/2019	
9058	2019	All other reportable incidents	47265	Dual-chamber implantable pacemaker, rate-responsive	CT141	Electrodes/Leads and associated devices	F1905	Device Revision or Replacement	A0705	Battery Problem		
					CT145	Cardiovascular devices	F1905	Device Revision or Replacement	A0705	Battery Problem		
					CT1684	Cardiac pacemakers and associated devices	F1905	Device Revision or Replacement	A0705	Battery Problem		
					CT1685	Cardiac pacemakers	F1905	Device Revision or Replacement	A0705	Battery Problem		
					CT233	Surgical	F1905	Device Revision or Replacement	A0705	Battery Problem		
					CT2406	Implantable	F1905	Device Revision or Replacement	A0705	Battery Problem		
					CT2407	Active implantable	F1905	Device Revision or Replacement	A0705	Battery Problem		
					CT257	Implantable cardiac pacemakers	F1905	Device Revision or Replacement	A0705	Battery Problem		
					CT2727	Active implant devices/ancillaries	F1905	Device Revision or Replacement	A0705	Battery Problem		
					CT2750	Implantable cardiac electrical stimulators and associated devices	F1905	Device Revision or Replacement	A0705	Battery Problem		
					CT2866	Implantable cardiac electrical stimulators	F1905	Device Revision or Replacement	A0705	Battery Problem		
					CT2985	Active	F1905	Device Revision or Replacement	A0705	Battery Problem		
					CT3001	Implantable metal	F1905	Device Revision or Replacement	A0705	Battery Problem		
					CT3002	Implantable synthetic polymer	F1905	Device Revision or Replacement	A0705	Battery Problem		
					CT3006	Implantable inorganic material	F1905	Device Revision or Replacement	A0705	Battery Problem		
					CT314	Electrically powered	F1905	Device Revision or Replacement	A0705	Battery Problem		
					CT321	Long-term surgical invasive	F1905	Device Revision or Replacement	A0705	Battery Problem		
					CT449	Generators and associated devices	F1905	Device Revision or Replacement	A0705	Battery Problem		
					CT981	Single use	F1905	Device Revision or Replacement	A0705	Battery Problem		
					CT983	Surgical invasive	F1905	Device Revision or Replacement	A0705	Battery Problem		
9059	2020	All other reportable incidents	35852	Single-chamber implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	13/03/2019	13/03/2019
									A071102	Failure to Interrogate	13/03/2019	13/03/2019
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	13/03/2019	13/03/2019
									A071102	Failure to Interrogate	13/03/2019	13/03/2019
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery	13/03/2019	13/03/2019
									A071102	Failure to Interrogate	13/03/2019	13/03/2019
9060	2019	All other reportable incidents	47265	Dual-chamber implantable pacemaker, rate-responsive	CT141	Electrodes/Leads and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
					CT145	Cardiovascular devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
					CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
					CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
					CT233	Surgical	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
					CT2406	Implantable	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
					CT2407	Active implantable	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
					CT2727	Active implant devices/ancillaries	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
					CT2750	Implantable cardiac electrical stimulators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
					CT2985	Active	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
					CT3001	Implantable metal	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
					CT3002	Implantable synthetic polymer	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
					CT3006	Implantable inorganic material	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
					CT314	Electrically powered	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
					CT321	Long-term surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
					CT449	Generators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
					CT981	Single use	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
					CT983	Surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
9061	2020	All other reportable incidents	17882	Physiologic-monitoring defibrillation system	CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A0713	Defibrillation/Stimulation Problem		
							F05	Delay to Treatment/ Therapy	A0713	Defibrillation/Stimulation Problem		
							F26	No Health Consequences or Impact	A0713	Defibrillation/Stimulation Problem		
9062	2019	All other reportable incidents	47265	Dual-chamber implantable pacemaker, rate-responsive	CT141	Electrodes/Leads and associated devices	F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	10/06/2019	
					CT145	Cardiovascular devices	F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	10/06/2019	
					CT1684	Cardiac pacemakers and associated devices	F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	10/06/2019	
					CT1685	Cardiac pacemakers	F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	10/06/2019	
					CT233	Surgical	F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	10/06/2019	
					CT2406	Implantable	F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	10/06/2019	
					CT2407	Active implantable	F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	10/06/2019	
					CT257	Implantable cardiac pacemakers	F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	10/06/2019	
					CT2727	Active implant devices/ancillaries	F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	10/06/2019	
					CT2750	Implantable cardiac electrical stimulators and associated devices	F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	10/06/2019	
					CT2866	Implantable cardiac electrical stimulators	F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	10/06/2019	
					CT2985	Active	F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	10/06/2019	
					CT3001	Implantable metal	F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	10/06/2019	
					CT3002	Implantable synthetic polymer	F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	10/06/2019	
					CT3006	Implantable inorganic material	F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	10/06/2019	
					CT314	Electrically powered	F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	10/06/2019	
					CT321	Long-term surgical invasive	F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	10/06/2019	
					CT449	Generators and associated devices	F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	10/06/2019	
					CT981	Single use	F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	10/06/2019	
					CT983	Surgical invasive	F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	10/06/2019	
9063	2019	All other reportable incidents	36241	Pacemaker/defibrillator lead	CT1684	Cardiac pacemakers and associated devices	F1905	Device Revision or Replacement	A010402	Migration	14/05/2019	
9064	2020	All other reportable incidents	17882	Physiologic-monitoring defibrillation system	CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A0709	Device Sensing Problem		
							F26	No Health Consequences or Impact	A0709	Device Sensing Problem		
9065	2020	All other reportable incidents	37265	Dual-chamber implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	01/04/2020	01/04/2020
							F1901	Additional Surgery	A070504	Premature Discharge of Battery	01/04/2020	01/04/2020
9066	2020	All other reportable incidents	47265	Dual-chamber implantable pacemaker, rate-responsive	CT141	Electrodes/Leads and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	07/04/2020	07/04/2020
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	07/04/2020	07/04/2020
					CT145	Cardiovascular devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	07/04/2020	07/04/2020
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	07/04/2020	07/04/2020
					CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	07/04/2020	07/04/2020
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	07/04/2020	07/04/2020
					CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	07/04/2020	07/04/2020
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	07/04/2020	07/04/2020
					CT233	Surgical	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	07/04/2020	07/04/2020
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	07/04/2020	07/04/2020

					CT2406	Implantable	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	07/04/2020	07/04/2020
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	07/04/2020	07/04/2020
					CT2407	Active implantable	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	07/04/2020	07/04/2020
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	07/04/2020	07/04/2020
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	07/04/2020	07/04/2020
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	07/04/2020	07/04/2020
					CT2727	Active implant devices/ancillaries	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	07/04/2020	07/04/2020
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	07/04/2020	07/04/2020
					CT2750	Implantable cardiac electrical stimulators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	07/04/2020	07/04/2020
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	07/04/2020	07/04/2020
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	07/04/2020	07/04/2020
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	07/04/2020	07/04/2020
					CT2985	Active	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	07/04/2020	07/04/2020
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	07/04/2020	07/04/2020
					CT3001	Implantable metal	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	07/04/2020	07/04/2020
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	07/04/2020	07/04/2020
					CT3002	Implantable synthetic polymer	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	07/04/2020	07/04/2020
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	07/04/2020	07/04/2020
					CT3006	Implantable inorganic material	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	07/04/2020	07/04/2020
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	07/04/2020	07/04/2020
					CT314	Electrically powered	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	07/04/2020	07/04/2020
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	07/04/2020	07/04/2020
					CT321	Long-term surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	07/04/2020	07/04/2020
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	07/04/2020	07/04/2020
					CT449	Generators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	07/04/2020	07/04/2020
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	07/04/2020	07/04/2020
					CT981	Single use	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	07/04/2020	07/04/2020
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	07/04/2020	07/04/2020
					CT983	Surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	07/04/2020	07/04/2020
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	07/04/2020	07/04/2020
9067	2020	All other reportable incidents	36241	Pacemaker/defibrillator lead	CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A0722	Impedance Problem	24/03/2020	24/03/2020
							F05	Delay to Treatment/ Therapy	A0722	Impedance Problem	24/03/2020	24/03/2020
							F14	Prolonged Episode of Care	A0722	Impedance Problem	24/03/2020	24/03/2020
							F19	Surgical Intervention	A0722	Impedance Problem	24/03/2020	24/03/2020
9068	2020	All other reportable incidents	36241	Pacemaker/defibrillator lead	CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A05	Mechanical Problem	24/03/2020	24/03/2020
									A0722	Impedance Problem	24/03/2020	24/03/2020
							F19	Surgical Intervention	A05	Mechanical Problem	24/03/2020	24/03/2020
									A0722	Impedance Problem	24/03/2020	24/03/2020
							F1905	Device Revision or Replacement	A05	Mechanical Problem	24/03/2020	24/03/2020
									A0722	Impedance Problem	24/03/2020	24/03/2020
9069	2020	All other reportable incidents	47267	Single-chamber implantable pacemaker, rate-responsive	CT141	Electrodes/Leads and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	01/11/2021	01/11/2021
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	01/11/2021	01/11/2021
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	01/11/2021	01/11/2021
					CT145	Cardiovascular devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	01/11/2021	01/11/2021
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	01/11/2021	01/11/2021
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	01/11/2021	01/11/2021
					CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	01/11/2021	01/11/2021
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	01/11/2021	01/11/2021
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	01/11/2021	01/11/2021
					CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	01/11/2021	01/11/2021
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	01/11/2021	01/11/2021
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	01/11/2021	01/11/2021
					CT233	Surgical	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	01/11/2021	01/11/2021
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	01/11/2021	01/11/2021
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	01/11/2021	01/11/2021
					CT2406	Implantable	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	01/11/2021	01/11/2021
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	01/11/2021	01/11/2021
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	01/11/2021	01/11/2021
					CT2407	Active implantable	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	01/11/2021	01/11/2021
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	01/11/2021	01/11/2021
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	01/11/2021	01/11/2021
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	01/11/2021	01/11/2021
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	01/11/2021	01/11/2021
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	01/11/2021	01/11/2021
					CT2727	Active implant devices/ancillaries	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	01/11/2021	01/11/2021
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	01/11/2021	01/11/2021
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	01/11/2021	01/11/2021
					CT2750	Implantable cardiac electrical stimulators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	01/11/2021	01/11/2021
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	01/11/2021	01/11/2021
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	01/11/2021	01/11/2021
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	01/11/2021	01/11/2021
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	01/11/2021	01/11/2021
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	01/11/2021	01/11/2021
					CT2985	Active	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	01/11/2021	01/11/2021
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	01/11/2021	01/11/2021
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	01/11/2021	01/11/2021
					CT3001	Implantable metal	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	01/11/2021	01/11/2021
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	01/11/2021	01/11/2021
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	01/11/2021	01/11/2021
					CT3002	Implantable synthetic polymer	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	01/11/2021	01/11/2021
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	01/11/2021	01/11/2021
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	01/11/2021	01/11/2021
					CT3006	Implantable inorganic material	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	01/11/2021	01/11/2021
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	01/11/2021	01/11/2021
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	01/11/2021	01/11/2021
					CT314	Electrically powered	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	01/11/2021	01/11/2021
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	01/11/2021	01/11/2021
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	01/11/2021	01/11/2021

					CT321	Long-term surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	01/11/2021	01/11/2021
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	01/11/2021	01/11/2021
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	01/11/2021	01/11/2021
					CT449	Generators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	01/11/2021	01/11/2021
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	01/11/2021	01/11/2021
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	01/11/2021	01/11/2021
					CT981	Single use	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	01/11/2021	01/11/2021
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	01/11/2021	01/11/2021
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	01/11/2021	01/11/2021
					CT983	Surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	01/11/2021	01/11/2021
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	01/11/2021	01/11/2021
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	01/11/2021	01/11/2021
9070	2020	All other reportable incidents	47267	Single-chamber implantable pacemaker, rate-responsive	CT141	Electrodes/Leads and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	19/03/2020	19/03/2020
							F19	Surgical Intervention	A070504	Premature Discharge of Battery	19/03/2020	19/03/2020
					CT145	Cardiovascular devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	19/03/2020	19/03/2020
							F19	Surgical Intervention	A070504	Premature Discharge of Battery	19/03/2020	19/03/2020
					CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	19/03/2020	19/03/2020
							F19	Surgical Intervention	A070504	Premature Discharge of Battery	19/03/2020	19/03/2020
					CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	19/03/2020	19/03/2020
							F19	Surgical Intervention	A070504	Premature Discharge of Battery	19/03/2020	19/03/2020
					CT233	Surgical	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	19/03/2020	19/03/2020
							F19	Surgical Intervention	A070504	Premature Discharge of Battery	19/03/2020	19/03/2020
					CT2406	Implantable	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	19/03/2020	19/03/2020
							F19	Surgical Intervention	A070504	Premature Discharge of Battery	19/03/2020	19/03/2020
					CT2407	Active implantable	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	19/03/2020	19/03/2020
							F19	Surgical Intervention	A070504	Premature Discharge of Battery	19/03/2020	19/03/2020
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	19/03/2020	19/03/2020
							F19	Surgical Intervention	A070504	Premature Discharge of Battery	19/03/2020	19/03/2020
					CT2727	Active implant devices/ancillaries	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	19/03/2020	19/03/2020
							F19	Surgical Intervention	A070504	Premature Discharge of Battery	19/03/2020	19/03/2020
					CT2750	Implantable cardiac electrical stimulators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	19/03/2020	19/03/2020
							F19	Surgical Intervention	A070504	Premature Discharge of Battery	19/03/2020	19/03/2020
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	19/03/2020	19/03/2020
							F19	Surgical Intervention	A070504	Premature Discharge of Battery	19/03/2020	19/03/2020
					CT2985	Active	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	19/03/2020	19/03/2020
							F19	Surgical Intervention	A070504	Premature Discharge of Battery	19/03/2020	19/03/2020
					CT3001	Implantable metal	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	19/03/2020	19/03/2020
							F19	Surgical Intervention	A070504	Premature Discharge of Battery	19/03/2020	19/03/2020
					CT3002	Implantable synthetic polymer	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	19/03/2020	19/03/2020
							F19	Surgical Intervention	A070504	Premature Discharge of Battery	19/03/2020	19/03/2020
					CT3006	Implantable inorganic material	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	19/03/2020	19/03/2020
							F19	Surgical Intervention	A070504	Premature Discharge of Battery	19/03/2020	19/03/2020
					CT314	Electrically powered	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	19/03/2020	19/03/2020
							F19	Surgical Intervention	A070504	Premature Discharge of Battery	19/03/2020	19/03/2020
					CT321	Long-term surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	19/03/2020	19/03/2020
							F19	Surgical Intervention	A070504	Premature Discharge of Battery	19/03/2020	19/03/2020
					CT449	Generators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	19/03/2020	19/03/2020
							F19	Surgical Intervention	A070504	Premature Discharge of Battery	19/03/2020	19/03/2020
					CT981	Single use	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	19/03/2020	19/03/2020
							F19	Surgical Intervention	A070504	Premature Discharge of Battery	19/03/2020	19/03/2020
					CT983	Surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	19/03/2020	19/03/2020
							F19	Surgical Intervention	A070504	Premature Discharge of Battery	19/03/2020	19/03/2020
9071	2020	All other reportable incidents	35852	Single-chamber implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
9072	2019	All other reportable incidents	35223	Endocardial/interventricular septal pacing lead	CT1684	Cardiac pacemakers and associated devices	F1905	Device Revision or Replacement	A072201	High Impedance		
9073	2019	All other reportable incidents	60789	Intracardiac pacemaker	CT130	Catheters and associated devices	F14	Prolonged Episode of Care	A0413	Material Separation	04/06/2019	
					CT145	Cardiovascular devices	F14	Prolonged Episode of Care	A0413	Material Separation	04/06/2019	
					CT1684	Cardiac pacemakers and associated devices	F14	Prolonged Episode of Care	A0413	Material Separation	04/06/2019	
					CT1685	Cardiac pacemakers	F14	Prolonged Episode of Care	A0413	Material Separation	04/06/2019	
					CT233	Surgical	F14	Prolonged Episode of Care	A0413	Material Separation	04/06/2019	
					CT2406	Implantable	F14	Prolonged Episode of Care	A0413	Material Separation	04/06/2019	
					CT2407	Active implantable	F14	Prolonged Episode of Care	A0413	Material Separation	04/06/2019	
					CT257	Implantable cardiac pacemakers	F14	Prolonged Episode of Care	A0413	Material Separation	04/06/2019	
					CT2727	Active implant devices/ancillaries	F14	Prolonged Episode of Care	A0413	Material Separation	04/06/2019	
					CT2750	Implantable cardiac electrical stimulators and associated devices	F14	Prolonged Episode of Care	A0413	Material Separation	04/06/2019	
					CT2818	Catheterization/Cannulation/Fluid path devices	F14	Prolonged Episode of Care	A0413	Material Separation	04/06/2019	
					CT2866	Implantable cardiac electrical stimulators	F14	Prolonged Episode of Care	A0413	Material Separation	04/06/2019	
					CT2985	Active	F14	Prolonged Episode of Care	A0413	Material Separation	04/06/2019	
					CT3001	Implantable metal	F14	Prolonged Episode of Care	A0413	Material Separation	04/06/2019	
					CT3002	Implantable synthetic polymer	F14	Prolonged Episode of Care	A0413	Material Separation	04/06/2019	
					CT3006	Implantable inorganic material	F14	Prolonged Episode of Care	A0413	Material Separation	04/06/2019	
					CT314	Electrically powered	F14	Prolonged Episode of Care	A0413	Material Separation	04/06/2019	
					CT321	Long-term surgical invasive	F14	Prolonged Episode of Care	A0413	Material Separation	04/06/2019	
					CT981	Single use	F14	Prolonged Episode of Care	A0413	Material Separation	04/06/2019	
					CT983	Surgical invasive	F14	Prolonged Episode of Care	A0413	Material Separation	04/06/2019	
9074	2019	All other reportable incidents	37265	Dual-chamber implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A0722	Impedance Problem		
9075	2019	All other reportable incidents	47265	Dual-chamber implantable pacemaker, rate-responsive	CT141	Electrodes/Leads and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070909	Oversensing		
					CT145	Cardiovascular devices	E2403	No Clinical Signs, Symptoms or Conditions	A070909	Oversensing		
					CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070909	Oversensing		
					CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A070909	Oversensing		
					CT233	Surgical	E2403	No Clinical Signs, Symptoms or Conditions	A070909	Oversensing		
					CT2406	Implantable	E2403	No Clinical Signs, Symptoms or Conditions	A070909	Oversensing		
					CT2407	Active implantable	E2403	No Clinical Signs, Symptoms or Conditions	A070909	Oversensing		
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A070909	Oversensing		
					CT2727	Active implant devices/ancillaries	E2403	No Clinical Signs, Symptoms or Conditions	A070909	Oversensing		
					CT2750	Implantable cardiac electrical stimulators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070909	Oversensing		

					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A070909	Oversensing		
					CT2985	Active	E2403	No Clinical Signs, Symptoms or Conditions	A070909	Oversensing		
					CT3001	Implantable metal	E2403	No Clinical Signs, Symptoms or Conditions	A070909	Oversensing		
					CT3002	Implantable synthetic polymer	E2403	No Clinical Signs, Symptoms or Conditions	A070909	Oversensing		
					CT3006	Implantable inorganic material	E2403	No Clinical Signs, Symptoms or Conditions	A070909	Oversensing		
					CT314	Electrically powered	E2403	No Clinical Signs, Symptoms or Conditions	A070909	Oversensing		
					CT321	Long-term surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A070909	Oversensing		
					CT449	Generators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070909	Oversensing		
					CT981	Single use	E2403	No Clinical Signs, Symptoms or Conditions	A070909	Oversensing		
					CT983	Surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A070909	Oversensing		
9076	2019	All other reportable incidents	36241	Pacemaker/defibrillator lead	CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A072201	High Impedance		
9077	2019	All other reportable incidents	35852	Single-chamber implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A070909	Oversensing		
9078	2019	All other reportable incidents	35852	Single-chamber implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E060102	Atrial Fibrillation	A071302	Inappropriate/Inadequate Shock/Stimulation	07/02/2021	07/02/2021
							F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A071302	Inappropriate/Inadequate Shock/Stimulation	07/02/2021	07/02/2021
							F26	No Health Consequences or Impact	A071302	Inappropriate/Inadequate Shock/Stimulation	07/02/2021	07/02/2021
9079	2019	All other reportable incidents	35852	Single-chamber implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A070910	Undersensing		
9080	2019	All other reportable incidents	36241	Pacemaker/defibrillator lead	CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A	Annex A		
9081	2020	All other reportable incidents	35223	Endocardial/interventricular septal pacing lead	CT1684	Cardiac pacemakers and associated devices	E2401	Insufficient Information	A070102	High Capture Threshold		
									A0722	Impedance Problem		
							E2403	No Clinical Signs, Symptoms or Conditions	A070102	High Capture Threshold		
									A0722	Impedance Problem		
							F26	No Health Consequences or Impact	A070102	High Capture Threshold		
									A0722	Impedance Problem		
9082	2019	Death	35852	Single-chamber implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A072201	High Impedance	17/05/2019	
							F21	Unexpected Deterioration	A072201	High Impedance	17/05/2019	
9083	2020	All other reportable incidents	47265	Dual-chamber implantable pacemaker, rate-responsive	CT141	Electrodes/Leads and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	22/04/2020	22/04/2020
							F19	Surgical Intervention	A070504	Premature Discharge of Battery	22/04/2020	22/04/2020
					CT145	Cardiovascular devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	22/04/2020	22/04/2020
							F19	Surgical Intervention	A070504	Premature Discharge of Battery	22/04/2020	22/04/2020
					CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	22/04/2020	22/04/2020
							F19	Surgical Intervention	A070504	Premature Discharge of Battery	22/04/2020	22/04/2020
					CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	22/04/2020	22/04/2020
							F19	Surgical Intervention	A070504	Premature Discharge of Battery	22/04/2020	22/04/2020
					CT233	Surgical	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	22/04/2020	22/04/2020
							F19	Surgical Intervention	A070504	Premature Discharge of Battery	22/04/2020	22/04/2020
					CT2406	Implantable	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	22/04/2020	22/04/2020
							F19	Surgical Intervention	A070504	Premature Discharge of Battery	22/04/2020	22/04/2020
					CT2407	Active implantable	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	22/04/2020	22/04/2020
							F19	Surgical Intervention	A070504	Premature Discharge of Battery	22/04/2020	22/04/2020
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	22/04/2020	22/04/2020
							F19	Surgical Intervention	A070504	Premature Discharge of Battery	22/04/2020	22/04/2020
					CT2727	Active implant devices/ancillaries	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	22/04/2020	22/04/2020
							F19	Surgical Intervention	A070504	Premature Discharge of Battery	22/04/2020	22/04/2020
					CT2750	Implantable cardiac electrical stimulators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	22/04/2020	22/04/2020
							F19	Surgical Intervention	A070504	Premature Discharge of Battery	22/04/2020	22/04/2020
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	22/04/2020	22/04/2020
							F19	Surgical Intervention	A070504	Premature Discharge of Battery	22/04/2020	22/04/2020
					CT2985	Active	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	22/04/2020	22/04/2020
							F19	Surgical Intervention	A070504	Premature Discharge of Battery	22/04/2020	22/04/2020
					CT3001	Implantable metal	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	22/04/2020	22/04/2020
							F19	Surgical Intervention	A070504	Premature Discharge of Battery	22/04/2020	22/04/2020
					CT3002	Implantable synthetic polymer	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	22/04/2020	22/04/2020
							F19	Surgical Intervention	A070504	Premature Discharge of Battery	22/04/2020	22/04/2020
					CT3006	Implantable inorganic material	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	22/04/2020	22/04/2020
							F19	Surgical Intervention	A070504	Premature Discharge of Battery	22/04/2020	22/04/2020
					CT314	Electrically powered	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	22/04/2020	22/04/2020
							F19	Surgical Intervention	A070504	Premature Discharge of Battery	22/04/2020	22/04/2020
					CT321	Long-term surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	22/04/2020	22/04/2020
							F19	Surgical Intervention	A070504	Premature Discharge of Battery	22/04/2020	22/04/2020
					CT449	Generators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	22/04/2020	22/04/2020
							F19	Surgical Intervention	A070504	Premature Discharge of Battery	22/04/2020	22/04/2020
					CT981	Single use	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	22/04/2020	22/04/2020
							F19	Surgical Intervention	A070504	Premature Discharge of Battery	22/04/2020	22/04/2020
					CT983	Surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	22/04/2020	22/04/2020
							F19	Surgical Intervention	A070504	Premature Discharge of Battery	22/04/2020	22/04/2020
9084	2019	Death	36241	Pacemaker/defibrillator lead	CT1684	Cardiac pacemakers and associated devices			A0721	Circuit Failure	17/05/2019	
9085	2020	All other reportable incidents	35852	Single-chamber implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A071302	Inappropriate/Inadequate Shock/Stimulation		
9086	2020	All other reportable incidents	35852	Single-chamber implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A070909	Oversensing		
9087	2020	All other reportable incidents	35852	Single-chamber implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A071302	Inappropriate/Inadequate Shock/Stimulation		
9088	2020	All other reportable incidents	17882	Physiologic-monitoring defibrillation system	CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070908	Failure to Sense		
							F26	No Health Consequences or Impact	A070908	Failure to Sense		
9089	2020	All other reportable incidents	17882	Physiologic-monitoring defibrillation system	CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070803	Failure to Power Up		
							F26	No Health Consequences or Impact	A070803	Failure to Power Up		
9090	2020	All other reportable incidents	17882	Physiologic-monitoring defibrillation system	CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070908	Failure to Sense		
							F26	No Health Consequences or Impact	A070908	Failure to Sense		
9091	2020	All other reportable incidents	17882	Physiologic-monitoring defibrillation system	CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070802	Intermittent Loss of Power		
							F27	Problem Identified During Non-Clinical Procedure	A070802	Intermittent Loss of Power		
9092	2019	All other reportable incidents	35852	Single-chamber implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A070909	Oversensing		
9093	2020	All other reportable incidents	37265	Dual-chamber implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A160601	Premature Elective Replacement Indicator		
							F1903	Device Explantation	A160601	Premature Elective Replacement Indicator		
9094	2020	All other reportable incidents	35223	Endocardial/interventricular septal pacing lead	CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A040101	Fracture		
									A0722	Impedance Problem		
							F26	No Health Consequences or Impact	A040101	Fracture		
									A0722	Impedance Problem		
9095	2020	All other reportable incidents	37265	Dual-chamber implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
9096	2020	All other reportable incidents	17882	Physiologic-monitoring defibrillation system	CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A071301	Failure to Deliver Shock/Stimulation		
							F26	No Health Consequences or Impact	A071301	Failure to Deliver Shock/Stimulation		
9097	2020	All other reportable incidents	17882	Physiologic-monitoring defibrillation system	CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A071301	Failure to Deliver Shock/Stimulation		

							F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A071301	Failure to Deliver Shock/Stimulation		
							F26	No Health Consequences or Impact	A071301	Failure to Deliver Shock/Stimulation		
9098	2020	All other reportable incidents	37265	Dual-chamber implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	09/04/2020	09/04/2020
							F1901	Additional Surgery	A070504	Premature Discharge of Battery	09/04/2020	09/04/2020
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	09/04/2020	09/04/2020
9099	2020	All other reportable incidents	37265	Dual-chamber implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A0203	Defective Device	28/01/2018	28/01/2018
									A072201	High Impedance	28/01/2018	28/01/2018
							F1905	Device Revision or Replacement	A0203	Defective Device	28/01/2018	28/01/2018
									A072201	High Impedance	28/01/2018	28/01/2018
							F24	Insufficient Health Impact Information	A0203	Defective Device	28/01/2018	28/01/2018
									A072201	High Impedance	28/01/2018	28/01/2018
9100	2020	All other reportable incidents	37265	Dual-chamber implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem	13/07/2020	13/07/2020
									A070504	Premature Discharge of Battery	13/07/2020	13/07/2020
							F1905	Device Revision or Replacement	A0705	Battery Problem	13/07/2020	13/07/2020
									A070504	Premature Discharge of Battery	13/07/2020	13/07/2020
9101	2020	All other reportable incidents	37265	Dual-chamber implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	F1905	Device Revision or Replacement	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
9102	2020	All other reportable incidents	37265	Dual-chamber implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	23/04/2020	23/04/2020
							F19	Surgical Intervention	A070504	Premature Discharge of Battery	23/04/2020	23/04/2020
9103	2020	All other reportable incidents	47265	Dual-chamber implantable pacemaker, rate-responsive	CT141	Electrodes/Leads and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
					CT145	Cardiovascular devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
					CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
					CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
					CT233	Surgical	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
					CT2406	Implantable	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
					CT2407	Active implantable	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
					CT2727	Active implant devices/ancillaries	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
					CT2750	Implantable cardiac electrical stimulators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
					CT2985	Active	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
					CT3001	Implantable metal	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
					CT3002	Implantable synthetic polymer	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
					CT3006	Implantable inorganic material	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
					CT314	Electrically powered	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
					CT321	Long-term surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
					CT449	Generators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
					CT981	Single use	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
					CT983	Surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
9104	2019	All other reportable incidents	36241	Pacemaker/defibrillator lead	CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A040101	Fracture	07/05/2019	
9105	2019	All other reportable incidents	36241	Pacemaker/defibrillator lead	CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070910	Undersensing	21/05/2019	
9106	2020	All other reportable incidents	35852	Single-chamber implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E2104	Electric Shock	A071302	Inappropriate/Inadequate Shock/Stimulation		
							F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A071302	Inappropriate/Inadequate Shock/Stimulation		
							F26	No Health Consequences or Impact	A071302	Inappropriate/Inadequate Shock/Stimulation		
9107	2019	All other reportable incidents	35852	Single-chamber implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A070909	Oversensing		
9108	2019	All other reportable incidents	36241	Pacemaker/defibrillator lead	CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A0510	Retraction Problem	02/04/2019	
9109	2019	All other reportable incidents	35852	Single-chamber implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A070909	Oversensing		
9110	2019	All other reportable incidents	36241	Pacemaker/defibrillator lead	CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070909	Oversensing		
9111	2020	All other reportable incidents	36241	Pacemaker/defibrillator lead	CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A0722	Impedance Problem	24/03/2020	24/03/2020
							F1905	Device Revision or Replacement	A0722	Impedance Problem	24/03/2020	24/03/2020
							F26	No Health Consequences or Impact	A0722	Impedance Problem	24/03/2020	24/03/2020
9112	2019	All other reportable incidents	35223	Endocardial/interventricular septal pacing lead	CT1684	Cardiac pacemakers and associated devices	F1905	Device Revision or Replacement	A0405	Degraded	30/05/2019	
9113	2020	All other reportable incidents	17882	Physiologic-monitoring defibrillation system	CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A07	Electrical /Electronic Property Problem		
									A070802	Intermittent Loss of Power		
							F27	Problem Identified During Non-Clinical Procedure	A07	Electrical /Electronic Property Problem		
									A070802	Intermittent Loss of Power		
9114	2019	All other reportable incidents	35852	Single-chamber implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A070909	Oversensing		
9115	2020	All other reportable incidents	17882	Physiologic-monitoring defibrillation system	CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A07	Electrical /Electronic Property Problem		
									A070802	Intermittent Loss of Power		
							F27	Problem Identified During Non-Clinical Procedure	A07	Electrical /Electronic Property Problem		
									A070802	Intermittent Loss of Power		
9116	2019	All other reportable incidents	36241	Pacemaker/defibrillator lead	CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070909	Oversensing		
9117	2019	All other reportable incidents	35852	Single-chamber implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A070909	Oversensing		
9118	2019	All other reportable incidents	36241	Pacemaker/defibrillator lead	CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070909	Oversensing		
9119	2019	All other reportable incidents	35852	Single-chamber implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A070909	Oversensing		
9120	2019	All other reportable incidents	36241	Pacemaker/defibrillator lead	CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A0510	Retraction Problem	21/05/2019	
9121	2019	All other reportable incidents	36241	Pacemaker/defibrillator lead	CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A05	Mechanical Problem	21/05/2019	21/05/2019
9122	2020	All other reportable incidents	47270	Cardiac resynchronization therapy implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	09/04/2020	09/04/2020
							F1901	Additional Surgery	A070504	Premature Discharge of Battery	09/04/2020	09/04/2020
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	09/04/2020	09/04/2020
9123	2020	All other reportable incidents	36241	Pacemaker/defibrillator lead	CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A05	Mechanical Problem	20/03/2020	20/03/2020
									A051201	Device Dislodged or Dislocated	20/03/2020	20/03/2020
									A0709	Device Sensing Problem	20/03/2020	20/03/2020
							F26	No Health Consequences or Impact	A05	Mechanical Problem	20/03/2020	20/03/2020
									A051201	Device Dislodged or Dislocated	20/03/2020	20/03/2020
									A0709	Device Sensing Problem	20/03/2020	20/03/2020
9124	2020	All other reportable incidents	47267	Single-chamber implantable pacemaker, rate-responsive	CT141	Electrodes/Leads and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT145	Cardiovascular devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT233	Surgical	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		

					CT2406	Implantable	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT2407	Active implantable	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT2727	Active implant devices/ancillaries	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT2750	Implantable cardiac electrical stimulators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT2985	Active	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT3001	Implantable metal	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT3002	Implantable synthetic polymer	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT3006	Implantable inorganic material	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT314	Electrically powered	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT321	Long-term surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT449	Generators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT981	Single use	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT983	Surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
9125	2020	All other reportable incidents	37265	Dual-chamber implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A070909	Oversensing		
9126	2019	All other reportable incidents	47265	Dual-chamber implantable pacemaker, rate-responsive	CT141	Electrodes/Leads and associated devices	F1905	Device Revision or Replacement	A160602	Premature End-of-Life Indicator	29/04/2019	
					CT145	Cardiovascular devices	F1905	Device Revision or Replacement	A160602	Premature End-of-Life Indicator	29/04/2019	
					CT1684	Cardiac pacemakers and associated devices	F1905	Device Revision or Replacement	A160602	Premature End-of-Life Indicator	29/04/2019	
					CT1685	Cardiac pacemakers	F1905	Device Revision or Replacement	A160602	Premature End-of-Life Indicator	29/04/2019	
					CT233	Surgical	F1905	Device Revision or Replacement	A160602	Premature End-of-Life Indicator	29/04/2019	
					CT2406	Implantable	F1905	Device Revision or Replacement	A160602	Premature End-of-Life Indicator	29/04/2019	
					CT2407	Active implantable	F1905	Device Revision or Replacement	A160602	Premature End-of-Life Indicator	29/04/2019	
					CT257	Implantable cardiac pacemakers	F1905	Device Revision or Replacement	A160602	Premature End-of-Life Indicator	29/04/2019	
					CT2727	Active implant devices/ancillaries	F1905	Device Revision or Replacement	A160602	Premature End-of-Life Indicator	29/04/2019	
					CT2750	Implantable cardiac electrical stimulators and associated devices	F1905	Device Revision or Replacement	A160602	Premature End-of-Life Indicator	29/04/2019	
					CT2866	Implantable cardiac electrical stimulators	F1905	Device Revision or Replacement	A160602	Premature End-of-Life Indicator	29/04/2019	
					CT2985	Active	F1905	Device Revision or Replacement	A160602	Premature End-of-Life Indicator	29/04/2019	
					CT3001	Implantable metal	F1905	Device Revision or Replacement	A160602	Premature End-of-Life Indicator	29/04/2019	
					CT3002	Implantable synthetic polymer	F1905	Device Revision or Replacement	A160602	Premature End-of-Life Indicator	29/04/2019	
					CT3006	Implantable inorganic material	F1905	Device Revision or Replacement	A160602	Premature End-of-Life Indicator	29/04/2019	
					CT314	Electrically powered	F1905	Device Revision or Replacement	A160602	Premature End-of-Life Indicator	29/04/2019	
					CT321	Long-term surgical invasive	F1905	Device Revision or Replacement	A160602	Premature End-of-Life Indicator	29/04/2019	
					CT449	Generators and associated devices	F1905	Device Revision or Replacement	A160602	Premature End-of-Life Indicator	29/04/2019	
					CT981	Single use	F1905	Device Revision or Replacement	A160602	Premature End-of-Life Indicator	29/04/2019	
					CT983	Surgical invasive	F1905	Device Revision or Replacement	A160602	Premature End-of-Life Indicator	29/04/2019	
9127	2019	All other reportable incidents	47270	Cardiac resynchronization therapy implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery		
9128	2020	All other reportable incidents	47270	Cardiac resynchronization therapy implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	21/04/2020	21/04/2020
							F1901	Additional Surgery	A070504	Premature Discharge of Battery	21/04/2020	21/04/2020
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	21/04/2020	21/04/2020
9129	2020	All other reportable incidents	35223	Endocardial/interventricular septal pacing lead	CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070909	Oversensing		
							F1901	Additional Surgery	A070909	Oversensing		
							F1905	Device Revision or Replacement	A070909	Oversensing		
9130	2019	All other reportable incidents	36241	Pacemaker/defibrillator lead	CT1684	Cardiac pacemakers and associated devices	E2008	Foreign Body In Patient	A0413	Material Separation	22/05/2019	
							F1905	Device Revision or Replacement	A0413	Material Separation	22/05/2019	
9131	2019	All other reportable incidents	35852	Single-chamber implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A070909	Oversensing		
9132	2019	All other reportable incidents	17882	Physiologic-monitoring defibrillation system	CT1684	Cardiac pacemakers and associated devices	F14	Prolonged Episode of Care	A070802	Intermittent Loss of Power		
9133	2019	All other reportable incidents	36241	Pacemaker/defibrillator lead	CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A072201	High Impedance		
9134	2019	All other reportable incidents	47265	Dual-chamber implantable pacemaker, rate-responsive	CT141	Electrodes/Leads and associated devices	F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A0701	Capturing Problem		
					CT145	Cardiovascular devices	F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A0701	Capturing Problem		
					CT1684	Cardiac pacemakers and associated devices	F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A0701	Capturing Problem		
					CT1685	Cardiac pacemakers	F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A0701	Capturing Problem		
					CT233	Surgical	F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A0701	Capturing Problem		
					CT2406	Implantable	F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A0701	Capturing Problem		
					CT2407	Active implantable	F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A0701	Capturing Problem		
					CT257	Implantable cardiac pacemakers	F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A0701	Capturing Problem		
					CT2727	Active implant devices/ancillaries	F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A0701	Capturing Problem		
					CT2750	Implantable cardiac electrical stimulators and associated devices	F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A0701	Capturing Problem		
					CT2866	Implantable cardiac electrical stimulators	F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A0701	Capturing Problem		
					CT2985	Active	F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A0701	Capturing Problem		
					CT3001	Implantable metal	F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A0701	Capturing Problem		
					CT3002	Implantable synthetic polymer	F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A0701	Capturing Problem		
					CT3006	Implantable inorganic material	F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A0701	Capturing Problem		
					CT314	Electrically powered	F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A0701	Capturing Problem		
					CT321	Long-term surgical invasive	F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A0701	Capturing Problem		
					CT449	Generators and associated devices	F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A0701	Capturing Problem		
					CT981	Single use	F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A0701	Capturing Problem		
					CT983	Surgical invasive	F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A0701	Capturing Problem		
9135	2020	All other reportable incidents	47267	Single-chamber implantable pacemaker, rate-responsive	CT141	Electrodes/Leads and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		

							F26	No Health Consequences or Impact	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
					CT145	Cardiovascular devices	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
					CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
					CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
					CT233	Surgical	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
					CT2406	Implantable	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
					CT2407	Active implantable	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
					CT2727	Active implant devices/ancillaries	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
					CT2750	Implantable cardiac electrical stimulators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
					CT2985	Active	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
					CT3001	Implantable metal	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
					CT3002	Implantable synthetic polymer	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
					CT3006	Implantable inorganic material	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
					CT314	Electrically powered	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
					CT321	Long-term surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
					CT449	Generators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
					CT981	Single use	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
					CT983	Surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
9136	2019	All other reportable incidents	35852	Single-chamber implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A072201	High Impedance		
9137	2019	All other reportable incidents	35852	Single-chamber implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A070909	Oversensing		
9138	2020	All other reportable incidents	35852	Single-chamber implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A071302	Inappropriate/Inadequate Shock/Stimulation		
9139	2020	All other reportable incidents	47265	Dual-chamber implantable pacemaker, rate-responsive	CT141	Electrodes/Leads and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
					CT145	Cardiovascular devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
					CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
					CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
					CT233	Surgical	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
					CT2406	Implantable	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
					CT2407	Active implantable	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		

					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
					CT2727	Active implant devices/ancillaries	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
					CT2750	Implantable cardiac electrical stimulators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
					CT2985	Active	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
					CT3001	Implantable metal	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
					CT3002	Implantable synthetic polymer	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
					CT3006	Implantable inorganic material	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
					CT314	Electrically powered	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
					CT321	Long-term surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
					CT449	Generators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
					CT981	Single use	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
					CT983	Surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
9140	2020	All other reportable incidents	35852	Single-chamber implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A070909	Oversensing		
9141	2019	All other reportable incidents	37265	Dual-chamber implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	17/06/2019	
9142	2019	All other reportable incidents	17882	Physiologic-monitoring defibrillation system	CT1684	Cardiac pacemakers and associated devices	F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A0709	Device Sensing Problem		
9143	2020	Death	35852	Single-chamber implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E060110	Ventricular Fibrillation	A070910	Undersensing	04/04/2020	04/04/2020
							F02	Death	A070910	Undersensing	04/04/2020	04/04/2020
							F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A070910	Undersensing	04/04/2020	04/04/2020
9144	2020	All other reportable incidents	35852	Single-chamber implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	23/10/2020	23/10/2020
									A070603	Failure to Charge	23/10/2020	23/10/2020
							F1901	Additional Surgery	A070504	Premature Discharge of Battery	23/10/2020	23/10/2020
									A070603	Failure to Charge	23/10/2020	23/10/2020
9145	2019	All other reportable incidents	35223	Endocardial/interventricular septal pacing lead	CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A072201	High Impedance		
9146	2020	All other reportable incidents	36241	Pacemaker/defibrillator lead	CT1684	Cardiac pacemakers and associated devices	F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A071302	Inappropriate/Inadequate Shock/Stimulation		
9147	2020	All other reportable incidents	35852	Single-chamber implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A070909	Oversensing		
9148	2020	All other reportable incidents	35852	Single-chamber implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A0908	Incorrect, Inadequate or Imprecise Result or Readings	03/04/2020	03/04/2020
							F19	Surgical Intervention	A0908	Incorrect, Inadequate or Imprecise Result or Readings	03/04/2020	03/04/2020
							F1905	Device Revision or Replacement	A0908	Incorrect, Inadequate or Imprecise Result or Readings	03/04/2020	03/04/2020
9149	2020	All other reportable incidents	37265	Dual-chamber implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	02/06/2020	02/06/2020
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery	02/06/2020	02/06/2020
9150	2019	All other reportable incidents	36241	Pacemaker/defibrillator lead	CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A072201	High Impedance		
9151	2019	All other reportable incidents	36241	Pacemaker/defibrillator lead	CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A072201	High Impedance		
9152	2020	All other reportable incidents	47270	Cardiac resynchronization therapy implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	16/04/2020	16/04/2020
							F1901	Additional Surgery	A070504	Premature Discharge of Battery	16/04/2020	16/04/2020
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	16/04/2020	16/04/2020
9153	2020	All other reportable incidents	47265	Dual-chamber implantable pacemaker, rate-responsive	CT141	Electrodes/Leads and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A071102	Failure to Interrogate	12/03/2020	12/03/2020
							F1901	Additional Surgery	A071102	Failure to Interrogate	12/03/2020	12/03/2020
							F1905	Device Revision or Replacement	A071102	Failure to Interrogate	12/03/2020	12/03/2020
					CT145	Cardiovascular devices	E2403	No Clinical Signs, Symptoms or Conditions	A071102	Failure to Interrogate	12/03/2020	12/03/2020
							F1901	Additional Surgery	A071102	Failure to Interrogate	12/03/2020	12/03/2020
							F1905	Device Revision or Replacement	A071102	Failure to Interrogate	12/03/2020	12/03/2020
					CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A071102	Failure to Interrogate	12/03/2020	12/03/2020
							F1901	Additional Surgery	A071102	Failure to Interrogate	12/03/2020	12/03/2020
							F1905	Device Revision or Replacement	A071102	Failure to Interrogate	12/03/2020	12/03/2020
					CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A071102	Failure to Interrogate	12/03/2020	12/03/2020
							F1901	Additional Surgery	A071102	Failure to Interrogate	12/03/2020	12/03/2020
							F1905	Device Revision or Replacement	A071102	Failure to Interrogate	12/03/2020	12/03/2020
					CT233	Surgical	E2403	No Clinical Signs, Symptoms or Conditions	A071102	Failure to Interrogate	12/03/2020	12/03/2020
							F1901	Additional Surgery	A071102	Failure to Interrogate	12/03/2020	12/03/2020
							F1905	Device Revision or Replacement	A071102	Failure to Interrogate	12/03/2020	12/03/2020
					CT2406	Implantable	E2403	No Clinical Signs, Symptoms or Conditions	A071102	Failure to Interrogate	12/03/2020	12/03/2020
							F1901	Additional Surgery	A071102	Failure to Interrogate	12/03/2020	12/03/2020
							F1905	Device Revision or Replacement	A071102	Failure to Interrogate	12/03/2020	12/03/2020
					CT2407	Active implantable	E2403	No Clinical Signs, Symptoms or Conditions	A071102	Failure to Interrogate	12/03/2020	12/03/2020
							F1901	Additional Surgery	A071102	Failure to Interrogate	12/03/2020	12/03/2020
							F1905	Device Revision or Replacement	A071102	Failure to Interrogate	12/03/2020	12/03/2020
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A071102	Failure to Interrogate	12/03/2020	12/03/2020
							F1901	Additional Surgery	A071102	Failure to Interrogate	12/03/2020	12/03/2020
							F1905	Device Revision or Replacement	A071102	Failure to Interrogate	12/03/2020	12/03/2020
					CT2727	Active implant devices/ancillaries	E2403	No Clinical Signs, Symptoms or Conditions	A071102	Failure to Interrogate	12/03/2020	12/03/2020
							F1901	Additional Surgery	A071102	Failure to Interrogate	12/03/2020	12/03/2020
							F1905	Device Revision or Replacement	A071102	Failure to Interrogate	12/03/2020	12/03/2020
					CT2750	Implantable cardiac electrical stimulators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A071102	Failure to Interrogate	12/03/2020	12/03/2020
							F1901	Additional Surgery	A071102	Failure to Interrogate	12/03/2020	12/03/2020
							F1905	Device Revision or Replacement	A071102	Failure to Interrogate	12/03/2020	12/03/2020
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A071102	Failure to Interrogate	12/03/2020	12/03/2020
							F1901	Additional Surgery	A071102	Failure to Interrogate	12/03/2020	12/03/2020
							F1905	Device Revision or Replacement	A071102	Failure to Interrogate	12/03/2020	12/03/2020
					CT2985	Active	E2403	No Clinical Signs, Symptoms or Conditions	A071102	Failure to Interrogate	12/03/2020	12/03/2020
							F1901	Additional Surgery	A071102	Failure to Interrogate	12/03/2020	12/03/2020
							F1905	Device Revision or Replacement	A071102	Failure to Interrogate	12/03/2020	12/03/2020
					CT3001	Implantable metal	E2403	No Clinical Signs, Symptoms or Conditions	A071102	Failure to Interrogate	12/03/2020	12/03/2020
							F1901	Additional Surgery	A071102	Failure to Interrogate	12/03/2020	12/03/2020
							F1905	Device Revision or Replacement	A071102	Failure to Interrogate	12/03/2020	12/03/2020
					CT3002	Implantable synthetic polymer	E2403	No Clinical Signs, Symptoms or Conditions	A071102	Failure to Interrogate	12/03/2020	12/03/2020
							F1901	Additional Surgery	A071102	Failure to Interrogate	12/03/2020	12/03/2020
							F1905	Device Revision or Replacement	A071102	Failure to Interrogate	12/03/2020	12/03/2020
					CT3006	Implantable inorganic material	E2403	No Clinical Signs, Symptoms or Conditions	A071102	Failure to Interrogate	12/03/2020	12/03/2020
							F1901	Additional Surgery	A071102	Failure to Interrogate	12/03/2020	12/03/2020
							F1905	Device Revision or Replacement	A071102	Failure to Interrogate	12/03/2020	12/03/2020
					CT314	Electrically powered	E2403	No Clinical Signs, Symptoms or Conditions	A071102	Failure to Interrogate	12/03/2020	12/03/2020
							F1901	Additional Surgery	A071102	Failure to Interrogate	12/03/2020	12/03/2020
							F1905	Device Revision or Replacement	A071102	Failure to Interrogate	12/03/2020	12/03/2020

					CT321	Long-term surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A071102	Failure to Interrogate	12/03/2020	12/03/2020
							F1901	Additional Surgery	A071102	Failure to Interrogate	12/03/2020	12/03/2020
							F1905	Device Revision or Replacement	A071102	Failure to Interrogate	12/03/2020	12/03/2020
					CT449	Generators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A071102	Failure to Interrogate	12/03/2020	12/03/2020
							F1901	Additional Surgery	A071102	Failure to Interrogate	12/03/2020	12/03/2020
							F1905	Device Revision or Replacement	A071102	Failure to Interrogate	12/03/2020	12/03/2020
					CT981	Single use	E2403	No Clinical Signs, Symptoms or Conditions	A071102	Failure to Interrogate	12/03/2020	12/03/2020
							F1901	Additional Surgery	A071102	Failure to Interrogate	12/03/2020	12/03/2020
							F1905	Device Revision or Replacement	A071102	Failure to Interrogate	12/03/2020	12/03/2020
					CT983	Surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A071102	Failure to Interrogate	12/03/2020	12/03/2020
							F1901	Additional Surgery	A071102	Failure to Interrogate	12/03/2020	12/03/2020
							F1905	Device Revision or Replacement	A071102	Failure to Interrogate	12/03/2020	12/03/2020
9154	2020	All other reportable incidents	36241	Pacemaker/defibrillator lead	CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070102	High Capture Threshold		
									A072201	High Impedance		
							F26	No Health Consequences or Impact	A070102	High Capture Threshold		
									A072201	High Impedance		
9155	2020	All other reportable incidents	35852	Single-chamber implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A070909	Oversensing		
									A0712	Pacing Problem		
									A072201	High Impedance		
							F1905	Device Revision or Replacement	A070909	Oversensing		
									A0712	Pacing Problem		
									A072201	High Impedance		
							F24	Insufficient Health Impact Information	A070909	Oversensing		
									A0712	Pacing Problem		
									A072201	High Impedance		
9156	2019	All other reportable incidents	47267	Single-chamber implantable pacemaker, rate-responsive	CT141	Electrodes/Leads and associated devices	F1905	Device Revision or Replacement	A071102	Failure to Interrogate	06/06/2019	
					CT145	Cardiovascular devices	F1905	Device Revision or Replacement	A071102	Failure to Interrogate	06/06/2019	
					CT1684	Cardiac pacemakers and associated devices	F1905	Device Revision or Replacement	A071102	Failure to Interrogate	06/06/2019	
					CT1685	Cardiac pacemakers	F1905	Device Revision or Replacement	A071102	Failure to Interrogate	06/06/2019	
					CT233	Surgical	F1905	Device Revision or Replacement	A071102	Failure to Interrogate	06/06/2019	
					CT2406	Implantable	F1905	Device Revision or Replacement	A071102	Failure to Interrogate	06/06/2019	
					CT2407	Active implantable	F1905	Device Revision or Replacement	A071102	Failure to Interrogate	06/06/2019	
					CT257	Implantable cardiac pacemakers	F1905	Device Revision or Replacement	A071102	Failure to Interrogate	06/06/2019	
					CT2727	Active implant devices/ancillaries	F1905	Device Revision or Replacement	A071102	Failure to Interrogate	06/06/2019	
					CT2750	Implantable cardiac electrical stimulators and associated devices	F1905	Device Revision or Replacement	A071102	Failure to Interrogate	06/06/2019	
					CT2866	Implantable cardiac electrical stimulators	F1905	Device Revision or Replacement	A071102	Failure to Interrogate	06/06/2019	
					CT2985	Active	F1905	Device Revision or Replacement	A071102	Failure to Interrogate	06/06/2019	
					CT3001	Implantable metal	F1905	Device Revision or Replacement	A071102	Failure to Interrogate	06/06/2019	
					CT3002	Implantable synthetic polymer	F1905	Device Revision or Replacement	A071102	Failure to Interrogate	06/06/2019	
					CT3006	Implantable inorganic material	F1905	Device Revision or Replacement	A071102	Failure to Interrogate	06/06/2019	
					CT314	Electrically powered	F1905	Device Revision or Replacement	A071102	Failure to Interrogate	06/06/2019	
					CT321	Long-term surgical invasive	F1905	Device Revision or Replacement	A071102	Failure to Interrogate	06/06/2019	
					CT449	Generators and associated devices	F1905	Device Revision or Replacement	A071102	Failure to Interrogate	06/06/2019	
					CT981	Single use	F1905	Device Revision or Replacement	A071102	Failure to Interrogate	06/06/2019	
					CT983	Surgical invasive	F1905	Device Revision or Replacement	A071102	Failure to Interrogate	06/06/2019	
9157	2019	All other reportable incidents	35852	Single-chamber implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A070909	Oversensing		
9158	2019	All other reportable incidents	35852	Single-chamber implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A072201	High Impedance	10/02/2021	10/02/2021
							F26	No Health Consequences or Impact	A072201	High Impedance	10/02/2021	10/02/2021
9159	2019	All other reportable incidents	35852	Single-chamber implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A072201	High Impedance		
9160	2020	All other reportable incidents	35852	Single-chamber implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A070909	Oversensing		
9161	2020	All other reportable incidents	47263	Cardiac resynchronization therapy implantable pacemaker	CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	31/03/2020	31/03/2020
							F12	Serious Injury/ Illness/ Impairment	A070504	Premature Discharge of Battery	31/03/2020	31/03/2020
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	31/03/2020	31/03/2020
					CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	31/03/2020	31/03/2020
							F12	Serious Injury/ Illness/ Impairment	A070504	Premature Discharge of Battery	31/03/2020	31/03/2020
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	31/03/2020	31/03/2020
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	31/03/2020	31/03/2020
							F12	Serious Injury/ Illness/ Impairment	A070504	Premature Discharge of Battery	31/03/2020	31/03/2020
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	31/03/2020	31/03/2020
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	31/03/2020	31/03/2020
							F12	Serious Injury/ Illness/ Impairment	A070504	Premature Discharge of Battery	31/03/2020	31/03/2020
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	31/03/2020	31/03/2020
9162	2020	All other reportable incidents	35852	Single-chamber implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A070909	Oversensing		
9163	2019	All other reportable incidents	47264	Dual-chamber implantable pacemaker, demand	CT141	Electrodes/Leads and associated devices	F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery		
					CT145	Cardiovascular devices	F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery		
					CT1684	Cardiac pacemakers and associated devices	F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery		
					CT1685	Cardiac pacemakers	F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery		
					CT233	Surgical	F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery		
					CT2406	Implantable	F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery		
					CT2407	Active implantable	F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery		
					CT257	Implantable cardiac pacemakers	F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery		
					CT2727	Active implant devices/ancillaries	F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery		
					CT2750	Implantable cardiac electrical stimulators and associated devices	F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery		
					CT2866	Implantable cardiac electrical stimulators	F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery		
					CT2985	Active	F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery		
					CT3001	Implantable metal	F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery		
					CT3002	Implantable synthetic polymer	F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery		
					CT3006	Implantable inorganic material	F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery		
					CT314	Electrically powered	F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery		
					CT321	Long-term surgical invasive	F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery		
					CT449	Generators and associated devices	F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery		
					CT981	Single use	F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery		
					CT983	Surgical invasive	F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery		
9164	2020	All other reportable incidents	47270	Cardiac resynchronization therapy implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A070909	Oversensing		
									A072202	Low Impedance		
									A1301	Failure to Read Input Signal		

							F26	No Health Consequences or Impact	A070909	Oversensing		
									A072202	Low Impedance		
									A1301	Failure to Read Input Signal		
9165	2020	All other reportable incidents	47265	Dual-chamber implantable pacemaker, rate-responsive	CT141	Electrodes/Leads and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070909	Oversensing		
									A0712	Pacing Problem		
									A072201	High Impedance		
							F21	Unexpected Deterioration	A070909	Oversensing		
									A0712	Pacing Problem		
									A072201	High Impedance		
							F26	No Health Consequences or Impact	A070909	Oversensing		
									A0712	Pacing Problem		
									A072201	High Impedance		
					CT145	Cardiovascular devices	E2403	No Clinical Signs, Symptoms or Conditions	A070909	Oversensing		
									A0712	Pacing Problem		
									A072201	High Impedance		
							F21	Unexpected Deterioration	A070909	Oversensing		
									A0712	Pacing Problem		
									A072201	High Impedance		
							F26	No Health Consequences or Impact	A070909	Oversensing		
									A0712	Pacing Problem		
									A072201	High Impedance		
					CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070909	Oversensing		
									A0712	Pacing Problem		
									A072201	High Impedance		
							F21	Unexpected Deterioration	A070909	Oversensing		
									A0712	Pacing Problem		
									A072201	High Impedance		
							F26	No Health Consequences or Impact	A070909	Oversensing		
									A0712	Pacing Problem		
									A072201	High Impedance		
					CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A070909	Oversensing		
									A0712	Pacing Problem		
									A072201	High Impedance		
							F21	Unexpected Deterioration	A070909	Oversensing		
									A0712	Pacing Problem		
									A072201	High Impedance		
							F26	No Health Consequences or Impact	A070909	Oversensing		
									A0712	Pacing Problem		
									A072201	High Impedance		
					CT233	Surgical	E2403	No Clinical Signs, Symptoms or Conditions	A070909	Oversensing		
									A0712	Pacing Problem		
									A072201	High Impedance		
							F21	Unexpected Deterioration	A070909	Oversensing		
									A0712	Pacing Problem		
									A072201	High Impedance		
							F26	No Health Consequences or Impact	A070909	Oversensing		
									A0712	Pacing Problem		
									A072201	High Impedance		
					CT2406	Implantable	E2403	No Clinical Signs, Symptoms or Conditions	A070909	Oversensing		
									A0712	Pacing Problem		
									A072201	High Impedance		
							F21	Unexpected Deterioration	A070909	Oversensing		
									A0712	Pacing Problem		
									A072201	High Impedance		
							F26	No Health Consequences or Impact	A070909	Oversensing		
									A0712	Pacing Problem		
									A072201	High Impedance		
					CT2407	Active implantable	E2403	No Clinical Signs, Symptoms or Conditions	A070909	Oversensing		
									A0712	Pacing Problem		
									A072201	High Impedance		
							F21	Unexpected Deterioration	A070909	Oversensing		
									A0712	Pacing Problem		
									A072201	High Impedance		
							F26	No Health Consequences or Impact	A070909	Oversensing		
									A0712	Pacing Problem		
									A072201	High Impedance		
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A070909	Oversensing		
									A0712	Pacing Problem		
									A072201	High Impedance		
							F21	Unexpected Deterioration	A070909	Oversensing		
									A0712	Pacing Problem		
									A072201	High Impedance		
							F26	No Health Consequences or Impact	A070909	Oversensing		
									A0712	Pacing Problem		
									A072201	High Impedance		
					CT2727	Active implant devices/ancillaries	E2403	No Clinical Signs, Symptoms or Conditions	A070909	Oversensing		
									A0712	Pacing Problem		
									A072201	High Impedance		
							F21	Unexpected Deterioration	A070909	Oversensing		
									A0712	Pacing Problem		
									A072201	High Impedance		
							F26	No Health Consequences or Impact	A070909	Oversensing		
									A0712	Pacing Problem		
									A072201	High Impedance		
					CT2750	Implantable cardiac electrical stimulators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070909	Oversensing		
									A0712	Pacing Problem		
									A072201	High Impedance		
							F21	Unexpected Deterioration	A070909	Oversensing		

									A0712	Pacing Problem		
									A072201	High Impedance		
							F26	No Health Consequences or Impact	A070909	Oversensing		
									A0712	Pacing Problem		
									A072201	High Impedance		
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A070909	Oversensing		
									A0712	Pacing Problem		
									A072201	High Impedance		
							F21	Unexpected Deterioration	A070909	Oversensing		
									A0712	Pacing Problem		
									A072201	High Impedance		
							F26	No Health Consequences or Impact	A070909	Oversensing		
									A0712	Pacing Problem		
									A072201	High Impedance		
					CT2985	Active	E2403	No Clinical Signs, Symptoms or Conditions	A070909	Oversensing		
									A0712	Pacing Problem		
									A072201	High Impedance		
							F21	Unexpected Deterioration	A070909	Oversensing		
									A0712	Pacing Problem		
									A072201	High Impedance		
							F26	No Health Consequences or Impact	A070909	Oversensing		
									A0712	Pacing Problem		
									A072201	High Impedance		
					CT3001	Implantable metal	E2403	No Clinical Signs, Symptoms or Conditions	A070909	Oversensing		
									A0712	Pacing Problem		
									A072201	High Impedance		
							F21	Unexpected Deterioration	A070909	Oversensing		
									A0712	Pacing Problem		
									A072201	High Impedance		
							F26	No Health Consequences or Impact	A070909	Oversensing		
									A0712	Pacing Problem		
									A072201	High Impedance		
					CT3002	Implantable synthetic polymer	E2403	No Clinical Signs, Symptoms or Conditions	A070909	Oversensing		
									A0712	Pacing Problem		
									A072201	High Impedance		
							F21	Unexpected Deterioration	A070909	Oversensing		
									A0712	Pacing Problem		
									A072201	High Impedance		
							F26	No Health Consequences or Impact	A070909	Oversensing		
									A0712	Pacing Problem		
									A072201	High Impedance		
					CT3006	Implantable inorganic material	E2403	No Clinical Signs, Symptoms or Conditions	A070909	Oversensing		
									A0712	Pacing Problem		
									A072201	High Impedance		
							F21	Unexpected Deterioration	A070909	Oversensing		
									A0712	Pacing Problem		
									A072201	High Impedance		
							F26	No Health Consequences or Impact	A070909	Oversensing		
									A0712	Pacing Problem		
									A072201	High Impedance		
					CT314	Electrically powered	E2403	No Clinical Signs, Symptoms or Conditions	A070909	Oversensing		
									A0712	Pacing Problem		
									A072201	High Impedance		
							F21	Unexpected Deterioration	A070909	Oversensing		
									A0712	Pacing Problem		
									A072201	High Impedance		
							F26	No Health Consequences or Impact	A070909	Oversensing		
									A0712	Pacing Problem		
									A072201	High Impedance		
					CT321	Long-term surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A070909	Oversensing		
									A0712	Pacing Problem		
									A072201	High Impedance		
							F21	Unexpected Deterioration	A070909	Oversensing		
									A0712	Pacing Problem		
									A072201	High Impedance		
							F26	No Health Consequences or Impact	A070909	Oversensing		
									A0712	Pacing Problem		
									A072201	High Impedance		
					CT449	Generators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070909	Oversensing		
									A0712	Pacing Problem		
									A072201	High Impedance		
							F21	Unexpected Deterioration	A070909	Oversensing		
									A0712	Pacing Problem		
									A072201	High Impedance		
							F26	No Health Consequences or Impact	A070909	Oversensing		
									A0712	Pacing Problem		
									A072201	High Impedance		
					CT981	Single use	E2403	No Clinical Signs, Symptoms or Conditions	A070909	Oversensing		
									A0712	Pacing Problem		
									A072201	High Impedance		
							F21	Unexpected Deterioration	A070909	Oversensing		
									A0712	Pacing Problem		
									A072201	High Impedance		
							F26	No Health Consequences or Impact	A070909	Oversensing		
									A0712	Pacing Problem		
									A072201	High Impedance		
					CT983	Surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A070909	Oversensing		
									A0712	Pacing Problem		

									A072201	High Impedance		
							F21	Unexpected Deterioration	A070909	Oversensing		
									A0712	Pacing Problem		
									A072201	High Impedance		
							F26	No Health Consequences or Impact	A070909	Oversensing		
									A0712	Pacing Problem		
									A072201	High Impedance		
9166	2020	All other reportable incidents	36241	Pacemaker/defibrillator lead	CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A0712	Pacing Problem		
9167	2020	All other reportable incidents	36241	Pacemaker/defibrillator lead	CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A05	Mechanical Problem	09/03/2020	09/03/2020
9168	2020	All other reportable incidents	35852	Single-chamber implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem	11/04/2020	11/04/2020
									A070504	Premature Discharge of Battery	11/04/2020	11/04/2020
							F08	Hospitalization or Prolonged Hospitalization	A0705	Battery Problem	11/04/2020	11/04/2020
									A070504	Premature Discharge of Battery	11/04/2020	11/04/2020
							F1905	Device Revision or Replacement	A0705	Battery Problem	11/04/2020	11/04/2020
									A070504	Premature Discharge of Battery	11/04/2020	11/04/2020
9169	2020	All other reportable incidents	37265	Dual-chamber implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	02/06/2020	02/06/2020
									A0908	Incorrect, Inadequate or Imprecise Result or Readings	02/06/2020	02/06/2020
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	02/06/2020	02/06/2020
									A0908	Incorrect, Inadequate or Imprecise Result or Readings	02/06/2020	02/06/2020
							F12	Serious Injury/ Illness/ Impairment	A070504	Premature Discharge of Battery	02/06/2020	02/06/2020
									A0908	Incorrect, Inadequate or Imprecise Result or Readings	02/06/2020	02/06/2020
							F19	Surgical Intervention	A070504	Premature Discharge of Battery	02/06/2020	02/06/2020
									A0908	Incorrect, Inadequate or Imprecise Result or Readings	02/06/2020	02/06/2020
9170	2019	All other reportable incidents	47265	Dual-chamber implantable pacemaker, rate-responsive	CT141	Electrodes/Leads and associated devices	F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	06/06/2019	
					CT145	Cardiovascular devices	F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	06/06/2019	
					CT1684	Cardiac pacemakers and associated devices	F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	06/06/2019	
					CT1685	Cardiac pacemakers	F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	06/06/2019	
					CT233	Surgical	F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	06/06/2019	
					CT2406	Implantable	F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	06/06/2019	
					CT2407	Active implantable	F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	06/06/2019	
					CT257	Implantable cardiac pacemakers	F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	06/06/2019	
					CT2727	Active implant devices/ancillaries	F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	06/06/2019	
					CT2750	Implantable cardiac electrical stimulators and associated devices	F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	06/06/2019	
					CT2866	Implantable cardiac electrical stimulators	F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	06/06/2019	
					CT2985	Active	F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	06/06/2019	
					CT3001	Implantable metal	F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	06/06/2019	
					CT3002	Implantable synthetic polymer	F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	06/06/2019	
					CT3006	Implantable inorganic material	F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	06/06/2019	
					CT314	Electrically powered	F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	06/06/2019	
					CT321	Long-term surgical invasive	F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	06/06/2019	
					CT449	Generators and associated devices	F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	06/06/2019	
					CT981	Single use	F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	06/06/2019	
					CT983	Surgical invasive	F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	06/06/2019	
9171	2020	All other reportable incidents	35852	Single-chamber implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A0909	Unexpected Therapeutic Results	01/03/2020	01/03/2020
							F08	Hospitalization or Prolonged Hospitalization	A0909	Unexpected Therapeutic Results	01/03/2020	01/03/2020
							F19	Surgical Intervention	A0909	Unexpected Therapeutic Results	01/03/2020	01/03/2020
							F1905	Device Revision or Replacement	A0909	Unexpected Therapeutic Results	01/03/2020	01/03/2020
9172	2020	All other reportable incidents	47265	Dual-chamber implantable pacemaker, rate-responsive	CT141	Electrodes/Leads and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
					CT145	Cardiovascular devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
					CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
					CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
					CT233	Surgical	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
					CT2406	Implantable	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
					CT2407	Active implantable	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
					CT2727	Active implant devices/ancillaries	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
					CT2750	Implantable cardiac electrical stimulators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
					CT2985	Active	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
					CT3001	Implantable metal	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
					CT3002	Implantable synthetic polymer	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
					CT3006	Implantable inorganic material	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
					CT314	Electrically powered	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
					CT321	Long-term surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
					CT449	Generators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
					CT981	Single use	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
					CT983	Surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
9173	2020	All other reportable incidents	37265	Dual-chamber implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	20/04/2020	20/04/2020
							F1901	Additional Surgery	A070504	Premature Discharge of Battery	20/04/2020	20/04/2020
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	20/04/2020	20/04/2020
9174	2020	All other reportable incidents	47270	Cardiac resynchronization therapy implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	24/04/2020	24/04/2020
							F1901	Additional Surgery	A070504	Premature Discharge of Battery	24/04/2020	24/04/2020
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	24/04/2020	24/04/2020
9175	2019	All other reportable incidents	37265	Dual-chamber implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	24/06/2019	
9176	2019	All other reportable incidents	37265	Dual-chamber implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A071302	Inappropriate/Inadequate Shock/Stimulation		
9177	2019	All other reportable incidents	36241	Pacemaker/defibrillator lead	CT1684	Cardiac pacemakers and associated devices	F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A071204	No Pacing		
9178	2020	All other reportable incidents	35852	Single-chamber implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	28/04/2020	28/04/2020
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	28/04/2020	28/04/2020
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery	28/04/2020	28/04/2020
9179	2019	All other reportable incidents	47265	Dual-chamber implantable pacemaker, rate-responsive	CT141	Electrodes/Leads and associated devices	F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	20/06/2019	
					CT145	Cardiovascular devices	F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	20/06/2019	
					CT1684	Cardiac pacemakers and associated devices	F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	20/06/2019	
					CT1685	Cardiac pacemakers	F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	20/06/2019	

					CT233	Surgical	F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	20/06/2019		
					CT2406	Implantable	F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	20/06/2019		
					CT2407	Active implantable	F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	20/06/2019		
					CT257	Implantable cardiac pacemakers	F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	20/06/2019		
					CT2727	Active implant devices/ancillaries	F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	20/06/2019		
					CT2750	Implantable cardiac electrical stimulators and associated devices	F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	20/06/2019		
					CT2866	Implantable cardiac electrical stimulators	F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	20/06/2019		
					CT2985	Active	F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	20/06/2019		
					CT3001	Implantable metal	F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	20/06/2019		
					CT3002	Implantable synthetic polymer	F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	20/06/2019		
					CT3006	Implantable inorganic material	F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	20/06/2019		
					CT314	Electrically powered	F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	20/06/2019		
					CT321	Long-term surgical invasive	F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	20/06/2019		
					CT449	Generators and associated devices	F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	20/06/2019		
					CT981	Single use	F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	20/06/2019		
					CT983	Surgical invasive	F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	20/06/2019		
9180	2020	All other reportable incidents	36241	Pacemaker/defibrillator lead	CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070910	Undersensing			
							F26	No Health Consequences or Impact	A070910	Undersensing			
9181	2020	All other reportable incidents	37265	Dual-chamber implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A090805	Incorrect Measurement	27/04/2020	27/04/2020	
							F1901	Additional Surgery	A090805	Incorrect Measurement	27/04/2020	27/04/2020	
							F1905	Device Revision or Replacement	A090805	Incorrect Measurement	27/04/2020	27/04/2020	
9182	2019	All other reportable incidents	47265	Dual-chamber implantable pacemaker, rate-responsive	CT141	Electrodes/Leads and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	13/05/2019		
					CT145	Cardiovascular devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	13/05/2019		
					CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	13/05/2019		
					CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	13/05/2019		
					CT233	Surgical	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	13/05/2019		
					CT2406	Implantable	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	13/05/2019		
					CT2407	Active implantable	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	13/05/2019		
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	13/05/2019		
					CT2727	Active implant devices/ancillaries	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	13/05/2019		
					CT2750	Implantable cardiac electrical stimulators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	13/05/2019		
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	13/05/2019		
					CT2985	Active	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	13/05/2019		
					CT3001	Implantable metal	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	13/05/2019		
					CT3002	Implantable synthetic polymer	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	13/05/2019		
					CT3006	Implantable inorganic material	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	13/05/2019		
					CT314	Electrically powered	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	13/05/2019		
					CT321	Long-term surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	13/05/2019		
					CT449	Generators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	13/05/2019		
					CT981	Single use	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	13/05/2019		
					CT983	Surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	13/05/2019		
9183	2019	All other reportable incidents	35852	Single-chamber implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery			
9184	2020	All other reportable incidents	35852	Single-chamber implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	17/04/2020	17/04/2020	
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	17/04/2020	17/04/2020	
							F12	Serious Injury/ Illness/ Impairment	A070504	Premature Discharge of Battery	17/04/2020	17/04/2020	
							F19	Surgical Intervention	A070504	Premature Discharge of Battery	17/04/2020	17/04/2020	
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	17/04/2020	17/04/2020	
9185	2020	All other reportable incidents	35852	Single-chamber implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	21/04/2020	21/04/2020	
							F1903	Device Explantation	A070504	Premature Discharge of Battery	21/04/2020	21/04/2020	
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	21/04/2020	21/04/2020	
9186	2020	All other reportable incidents	47270	Cardiac resynchronization therapy implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A04	Material Integrity Problem	13/04/2020	13/04/2020	
							F08	Hospitalization or Prolonged Hospitalization	A04	Material Integrity Problem	13/04/2020	13/04/2020	
							F1903	Device Explantation	A04	Material Integrity Problem	13/04/2020	13/04/2020	
9187	2019	All other reportable incidents	35852	Single-chamber implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A071302	Inappropriate/Inadequate Shock/Stimulation			
9188	2019	All other reportable incidents	35852	Single-chamber implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A071302	Inappropriate/Inadequate Shock/Stimulation			
9189	2019	All other reportable incidents	36241	Pacemaker/defibrillator lead	CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070909	Oversensing			
9190	2019	All other reportable incidents	35852	Single-chamber implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A071302	Inappropriate/Inadequate Shock/Stimulation			
9191	2019	All other reportable incidents	35852	Single-chamber implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	01/08/2019		
9192	2019	All other reportable incidents	36241	Pacemaker/defibrillator lead	CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A010402	Migration			
9193	2020	Death	37265	Dual-chamber implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A070909	Oversensing	24/03/2020	24/03/2020	
									A070910	Undersensing	24/03/2020	24/03/2020	
							F02	Death	A070909	Oversensing	24/03/2020	24/03/2020	
									A070910	Undersensing	24/03/2020	24/03/2020	
9194	2020	All other reportable incidents	36241	Pacemaker/defibrillator lead	CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070909	Oversensing			
9195	2020	All other reportable incidents	36241	Pacemaker/defibrillator lead	CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A072201	High Impedance			
							F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A072201	High Impedance			
							F26	No Health Consequences or Impact	A072201	High Impedance			
9196	2020	All other reportable incidents	37265	Dual-chamber implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A072201	High Impedance			
9197	2019	All other reportable incidents	36241	Pacemaker/defibrillator lead	CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070101	Failure to Capture			
									A072201	High Impedance			
							F08	Hospitalization or Prolonged Hospitalization	A070101	Failure to Capture			
									A072201	High Impedance			
							F1905	Device Revision or Replacement	A070101	Failure to Capture			
									A072201	High Impedance			
9198	2020	All other reportable incidents	47263	Cardiac resynchronization therapy implantable pacemaker	CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	22/05/2020	22/05/2020	
							F1901	Additional Surgery	A070504	Premature Discharge of Battery	22/05/2020	22/05/2020	
					CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	22/05/2020	22/05/2020	
							F1901	Additional Surgery	A070504	Premature Discharge of Battery	22/05/2020	22/05/2020	
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	22/05/2020	22/05/2020	
							F1901	Additional Surgery	A070504	Premature Discharge of Battery	22/05/2020	22/05/2020	
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	22/05/2020	22/05/2020	
							F1901	Additional Surgery	A070504	Premature Discharge of Battery	22/05/2020	22/05/2020	
9199	2019	All other reportable incidents	47270	Cardiac resynchronization therapy implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	25/06/2019		
9200	2019	All other reportable incidents	35852	Single-chamber implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E2104	Electric Shock	A071302	Inappropriate/Inadequate Shock/Stimulation			
9201	2020	All other reportable incidents	37265	Dual-chamber implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	12/05/2020	12/05/2020	

							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	12/05/2020	12/05/2020
							F12	Serious Injury/ Illness/ Impairment	A070504	Premature Discharge of Battery	12/05/2020	12/05/2020
							F1903	Device Explantation	A070504	Premature Discharge of Battery	12/05/2020	12/05/2020
9202	2020	All other reportable incidents	35223	Endocardial/interventricular septal pacing lead	CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A040101	Fracture		
									A090801	Signal Artifact/Noise		
							F26	No Health Consequences or Impact	A040101	Fracture		
									A090801	Signal Artifact/Noise		
9203	2020	All other reportable incidents	37265	Dual-chamber implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A160602	Premature End-of-Life Indicator	19/04/2020	19/04/2020
							F1905	Device Revision or Replacement	A160602	Premature End-of-Life Indicator	19/04/2020	19/04/2020
9204	2020	All other reportable incidents	17882	Physiologic-monitoring defibrillation system	CT1684	Cardiac pacemakers and associated devices	E060101	Asystole	A070802	Intermittent Loss of Power		
							E2401	Insufficient Information	A070802	Intermittent Loss of Power		
							F26	No Health Consequences or Impact	A070802	Intermittent Loss of Power		
9205	2020	All other reportable incidents	17882	Physiologic-monitoring defibrillation system	CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070908	Failure to Sense		
							F26	No Health Consequences or Impact	A070908	Failure to Sense		
9206	2019	All other reportable incidents	35223	Endocardial/interventricular septal pacing lead	CT1684	Cardiac pacemakers and associated devices	F1905	Device Revision or Replacement	A072201	High Impedance		
9207	2019	All other reportable incidents	47265	Dual-chamber implantable pacemaker, rate-responsive	CT141	Electrodes/Leads and associated devices	F1905	Device Revision or Replacement	A0712	Pacing Problem	13/06/2019	
					CT145	Cardiovascular devices	F1905	Device Revision or Replacement	A0712	Pacing Problem	13/06/2019	
					CT1684	Cardiac pacemakers and associated devices	F1905	Device Revision or Replacement	A0712	Pacing Problem	13/06/2019	
					CT1685	Cardiac pacemakers	F1905	Device Revision or Replacement	A0712	Pacing Problem	13/06/2019	
					CT233	Surgical	F1905	Device Revision or Replacement	A0712	Pacing Problem	13/06/2019	
					CT2406	Implantable	F1905	Device Revision or Replacement	A0712	Pacing Problem	13/06/2019	
					CT2407	Active implantable	F1905	Device Revision or Replacement	A0712	Pacing Problem	13/06/2019	
					CT257	Implantable cardiac pacemakers	F1905	Device Revision or Replacement	A0712	Pacing Problem	13/06/2019	
					CT2727	Active implant devices/ancillaries	F1905	Device Revision or Replacement	A0712	Pacing Problem	13/06/2019	
					CT2750	Implantable cardiac electrical stimulators and associated devices	F1905	Device Revision or Replacement	A0712	Pacing Problem	13/06/2019	
					CT2866	Implantable cardiac electrical stimulators	F1905	Device Revision or Replacement	A0712	Pacing Problem	13/06/2019	
					CT2985	Active	F1905	Device Revision or Replacement	A0712	Pacing Problem	13/06/2019	
					CT3001	Implantable metal	F1905	Device Revision or Replacement	A0712	Pacing Problem	13/06/2019	
					CT3002	Implantable synthetic polymer	F1905	Device Revision or Replacement	A0712	Pacing Problem	13/06/2019	
					CT3006	Implantable inorganic material	F1905	Device Revision or Replacement	A0712	Pacing Problem	13/06/2019	
					CT314	Electrically powered	F1905	Device Revision or Replacement	A0712	Pacing Problem	13/06/2019	
					CT321	Long-term surgical invasive	F1905	Device Revision or Replacement	A0712	Pacing Problem	13/06/2019	
					CT449	Generators and associated devices	F1905	Device Revision or Replacement	A0712	Pacing Problem	13/06/2019	
					CT981	Single use	F1905	Device Revision or Replacement	A0712	Pacing Problem	13/06/2019	
					CT983	Surgical invasive	F1905	Device Revision or Replacement	A0712	Pacing Problem	13/06/2019	
9208	2019	All other reportable incidents	47265	Dual-chamber implantable pacemaker, rate-responsive	CT141	Electrodes/Leads and associated devices	F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	18/06/2019	
					CT145	Cardiovascular devices	F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	18/06/2019	
					CT1684	Cardiac pacemakers and associated devices	F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	18/06/2019	
					CT1685	Cardiac pacemakers	F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	18/06/2019	
					CT233	Surgical	F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	18/06/2019	
					CT2406	Implantable	F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	18/06/2019	
					CT2407	Active implantable	F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	18/06/2019	
					CT257	Implantable cardiac pacemakers	F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	18/06/2019	
					CT2727	Active implant devices/ancillaries	F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	18/06/2019	
					CT2750	Implantable cardiac electrical stimulators and associated devices	F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	18/06/2019	
					CT2866	Implantable cardiac electrical stimulators	F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	18/06/2019	
					CT2985	Active	F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	18/06/2019	
					CT3001	Implantable metal	F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	18/06/2019	
					CT3002	Implantable synthetic polymer	F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	18/06/2019	
					CT3006	Implantable inorganic material	F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	18/06/2019	
					CT314	Electrically powered	F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	18/06/2019	
					CT321	Long-term surgical invasive	F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	18/06/2019	
					CT449	Generators and associated devices	F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	18/06/2019	
					CT981	Single use	F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	18/06/2019	
					CT983	Surgical invasive	F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	18/06/2019	
9209	2019	All other reportable incidents	35852	Single-chamber implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A070909	Oversensing	08/01/2020	08/01/2020
									A071302	Inappropriate/Inadequate Shock/Stimulation	08/01/2020	08/01/2020
9210	2019	All other reportable incidents	17882	Physiologic-monitoring defibrillation system	CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070803	Failure to Power Up		
							F27	Problem Identified During Non-Clinical Procedure	A070803	Failure to Power Up		
9211	2019	All other reportable incidents	17882	Physiologic-monitoring defibrillation system	CT1684	Cardiac pacemakers and associated devices	F05	Delay to Treatment/ Therapy	A071301	Failure to Deliver Shock/Stimulation		
9212	2019	All other reportable incidents	17882	Physiologic-monitoring defibrillation system	CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A	Annex A		
9213	2019	All other reportable incidents	36241	Pacemaker/defibrillator lead	CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A010402	Migration	06/06/2019	
9214	2020	All other reportable incidents	47265	Dual-chamber implantable pacemaker, rate-responsive	CT141	Electrodes/Leads and associated devices	E0112	Dizziness	A27	Appropriate Device Problem Term/Code not Available		
							E011903	Syncope/Fainting	A27	Appropriate Device Problem Term/Code not Available		
							F12	Serious Injury/ Illness/ Impairment	A27	Appropriate Device Problem Term/Code not Available		
							F21	Unexpected Deterioration	A27	Appropriate Device Problem Term/Code not Available		
					CT145	Cardiovascular devices	E0112	Dizziness	A27	Appropriate Device Problem Term/Code not Available		
							E011903	Syncope/Fainting	A27	Appropriate Device Problem Term/Code not Available		
							F12	Serious Injury/ Illness/ Impairment	A27	Appropriate Device Problem Term/Code not Available		
							F21	Unexpected Deterioration	A27	Appropriate Device Problem Term/Code not Available		
					CT1684	Cardiac pacemakers and associated devices	E0112	Dizziness	A27	Appropriate Device Problem Term/Code not Available		
							E011903	Syncope/Fainting	A27	Appropriate Device Problem Term/Code not Available		
							F12	Serious Injury/ Illness/ Impairment	A27	Appropriate Device Problem Term/Code not Available		
							F21	Unexpected Deterioration	A27	Appropriate Device Problem Term/Code not Available		
					CT1685	Cardiac pacemakers	E0112	Dizziness	A27	Appropriate Device Problem Term/Code not Available		

							E011903	Syncope/Fainting	A27	Appropriate Device Problem Term/Code not Available		
							F12	Serious Injury/ Illness/ Impairment	A27	Appropriate Device Problem Term/Code not Available		
							F21	Unexpected Deterioration	A27	Appropriate Device Problem Term/Code not Available		
					CT233	Surgical	E0112	Dizziness	A27	Appropriate Device Problem Term/Code not Available		
							E011903	Syncope/Fainting	A27	Appropriate Device Problem Term/Code not Available		
							F12	Serious Injury/ Illness/ Impairment	A27	Appropriate Device Problem Term/Code not Available		
							F21	Unexpected Deterioration	A27	Appropriate Device Problem Term/Code not Available		
					CT2406	Implantable	E0112	Dizziness	A27	Appropriate Device Problem Term/Code not Available		
							E011903	Syncope/Fainting	A27	Appropriate Device Problem Term/Code not Available		
							F12	Serious Injury/ Illness/ Impairment	A27	Appropriate Device Problem Term/Code not Available		
							F21	Unexpected Deterioration	A27	Appropriate Device Problem Term/Code not Available		
					CT2407	Active implantable	E0112	Dizziness	A27	Appropriate Device Problem Term/Code not Available		
							E011903	Syncope/Fainting	A27	Appropriate Device Problem Term/Code not Available		
							F12	Serious Injury/ Illness/ Impairment	A27	Appropriate Device Problem Term/Code not Available		
							F21	Unexpected Deterioration	A27	Appropriate Device Problem Term/Code not Available		
					CT257	Implantable cardiac pacemakers	E0112	Dizziness	A27	Appropriate Device Problem Term/Code not Available		
							E011903	Syncope/Fainting	A27	Appropriate Device Problem Term/Code not Available		
							F12	Serious Injury/ Illness/ Impairment	A27	Appropriate Device Problem Term/Code not Available		
							F21	Unexpected Deterioration	A27	Appropriate Device Problem Term/Code not Available		
					CT2727	Active implant devices/ancillaries	E0112	Dizziness	A27	Appropriate Device Problem Term/Code not Available		
							E011903	Syncope/Fainting	A27	Appropriate Device Problem Term/Code not Available		
							F12	Serious Injury/ Illness/ Impairment	A27	Appropriate Device Problem Term/Code not Available		
							F21	Unexpected Deterioration	A27	Appropriate Device Problem Term/Code not Available		
					CT2750	Implantable cardiac electrical stimulators and associated devices	E0112	Dizziness	A27	Appropriate Device Problem Term/Code not Available		
							E011903	Syncope/Fainting	A27	Appropriate Device Problem Term/Code not Available		
							F12	Serious Injury/ Illness/ Impairment	A27	Appropriate Device Problem Term/Code not Available		
							F21	Unexpected Deterioration	A27	Appropriate Device Problem Term/Code not Available		
					CT2866	Implantable cardiac electrical stimulators	E0112	Dizziness	A27	Appropriate Device Problem Term/Code not Available		
							E011903	Syncope/Fainting	A27	Appropriate Device Problem Term/Code not Available		
							F12	Serious Injury/ Illness/ Impairment	A27	Appropriate Device Problem Term/Code not Available		
							F21	Unexpected Deterioration	A27	Appropriate Device Problem Term/Code not Available		
					CT2985	Active	E0112	Dizziness	A27	Appropriate Device Problem Term/Code not Available		
							E011903	Syncope/Fainting	A27	Appropriate Device Problem Term/Code not Available		
							F12	Serious Injury/ Illness/ Impairment	A27	Appropriate Device Problem Term/Code not Available		
							F21	Unexpected Deterioration	A27	Appropriate Device Problem Term/Code not Available		
					CT3001	Implantable metal	E0112	Dizziness	A27	Appropriate Device Problem Term/Code not Available		
							E011903	Syncope/Fainting	A27	Appropriate Device Problem Term/Code not Available		
							F12	Serious Injury/ Illness/ Impairment	A27	Appropriate Device Problem Term/Code not Available		
							F21	Unexpected Deterioration	A27	Appropriate Device Problem Term/Code not Available		
					CT3002	Implantable synthetic polymer	E0112	Dizziness	A27	Appropriate Device Problem Term/Code not Available		
							E011903	Syncope/Fainting	A27	Appropriate Device Problem Term/Code not Available		
							F12	Serious Injury/ Illness/ Impairment	A27	Appropriate Device Problem Term/Code not Available		
							F21	Unexpected Deterioration	A27	Appropriate Device Problem Term/Code not Available		
					CT3006	Implantable inorganic material	E0112	Dizziness	A27	Appropriate Device Problem Term/Code not Available		
							E011903	Syncope/Fainting	A27	Appropriate Device Problem Term/Code not Available		
							F12	Serious Injury/ Illness/ Impairment	A27	Appropriate Device Problem Term/Code not Available		
							F21	Unexpected Deterioration	A27	Appropriate Device Problem Term/Code not Available		
					CT314	Electrically powered	E0112	Dizziness	A27	Appropriate Device Problem Term/Code not Available		
							E011903	Syncope/Fainting	A27	Appropriate Device Problem Term/Code not Available		
							F12	Serious Injury/ Illness/ Impairment	A27	Appropriate Device Problem Term/Code not Available		

							F21	Unexpected Deterioration	A27	Appropriate Device Problem Term/Code not Available		
					CT321	Long-term surgical invasive	E0112	Dizziness	A27	Appropriate Device Problem Term/Code not Available		
							E011903	Syncope/Fainting	A27	Appropriate Device Problem Term/Code not Available		
							F12	Serious Injury/ Illness/ Impairment	A27	Appropriate Device Problem Term/Code not Available		
							F21	Unexpected Deterioration	A27	Appropriate Device Problem Term/Code not Available		
					CT449	Generators and associated devices	E0112	Dizziness	A27	Appropriate Device Problem Term/Code not Available		
							E011903	Syncope/Fainting	A27	Appropriate Device Problem Term/Code not Available		
							F12	Serious Injury/ Illness/ Impairment	A27	Appropriate Device Problem Term/Code not Available		
							F21	Unexpected Deterioration	A27	Appropriate Device Problem Term/Code not Available		
					CT981	Single use	E0112	Dizziness	A27	Appropriate Device Problem Term/Code not Available		
							E011903	Syncope/Fainting	A27	Appropriate Device Problem Term/Code not Available		
							F12	Serious Injury/ Illness/ Impairment	A27	Appropriate Device Problem Term/Code not Available		
							F21	Unexpected Deterioration	A27	Appropriate Device Problem Term/Code not Available		
					CT983	Surgical invasive	E0112	Dizziness	A27	Appropriate Device Problem Term/Code not Available		
							E011903	Syncope/Fainting	A27	Appropriate Device Problem Term/Code not Available		
							F12	Serious Injury/ Illness/ Impairment	A27	Appropriate Device Problem Term/Code not Available		
							F21	Unexpected Deterioration	A27	Appropriate Device Problem Term/Code not Available		
9215	2020	All other reportable incidents	35852	Single-chamber implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A071302	Inappropriate/Inadequate Shock/Stimulation	15/06/2020	15/06/2020
							F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A071302	Inappropriate/Inadequate Shock/Stimulation	15/06/2020	15/06/2020
							F19	Surgical Intervention	A071302	Inappropriate/Inadequate Shock/Stimulation	15/06/2020	15/06/2020
9216	2020	All other reportable incidents	17882	Physiologic-monitoring defibrillation system	CT1684	Cardiac pacemakers and associated devices	E0602	Cardiac Arrest	A090402	Failure to Deliver Energy		
							F05	Delay to Treatment/ Therapy	A090402	Failure to Deliver Energy		
							F1001	Absence of Treatment	A090402	Failure to Deliver Energy		
9217	2020	All other reportable incidents	36241	Pacemaker/defibrillator lead	CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A072201	High Impedance		
							F26	No Health Consequences or Impact	A072201	High Impedance		
9218	2020	All other reportable incidents	47265	Dual-chamber implantable pacemaker, rate-responsive	CT141	Electrodes/Leads and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem		
					CT145	Cardiovascular devices	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem		
					CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem		
					CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem		
					CT233	Surgical	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem		
					CT2406	Implantable	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem		
					CT2407	Active implantable	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem		
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem		
					CT2727	Active implant devices/ancillaries	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem		
					CT2750	Implantable cardiac electrical stimulators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem		
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem		
					CT2985	Active	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem		
					CT3001	Implantable metal	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem		
					CT3002	Implantable synthetic polymer	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem		
					CT3006	Implantable inorganic material	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem		
					CT314	Electrically powered	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem		
					CT321	Long-term surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem		
					CT449	Generators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem		
					CT981	Single use	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem		
					CT983	Surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem		
9219	2019	All other reportable incidents	35852	Single-chamber implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A160601	Premature Elective Replacement Indicator		
9220	2019	All other reportable incidents	37265	Dual-chamber implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A070909	Oversensing		
9221	2020	All other reportable incidents	36241	Pacemaker/defibrillator lead	CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A072201	High Impedance		
							F26	No Health Consequences or Impact	A072201	High Impedance		
9222	2019	All other reportable incidents	12913	Pacemaker, implantable	CT145	Cardiovascular devices	F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A150101	Activation Failure	07/06/2019	
					CT1684	Cardiac pacemakers and associated devices	F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A150101	Activation Failure	07/06/2019	
					CT1685	Cardiac pacemakers	F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A150101	Activation Failure	07/06/2019	
					CT2406	Implantable	F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A150101	Activation Failure	07/06/2019	
					CT2407	Active implantable	F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A150101	Activation Failure	07/06/2019	
					CT257	Implantable cardiac pacemakers	F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A150101	Activation Failure	07/06/2019	
					CT2727	Active implant devices/ancillaries	F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A150101	Activation Failure	07/06/2019	
					CT2750	Implantable cardiac electrical stimulators and associated devices	F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A150101	Activation Failure	07/06/2019	
					CT2866	Implantable cardiac electrical stimulators	F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A150101	Activation Failure	07/06/2019	
					CT2985	Active	F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A150101	Activation Failure	07/06/2019	
					CT314	Electrically powered	F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A150101	Activation Failure	07/06/2019	
9223	2019	All other reportable incidents	47263	Cardiac resynchronization therapy implantable pacemaker	CT1684	Cardiac pacemakers and associated devices	F1905	Device Revision or Replacement	A160601	Premature Elective Replacement Indicator	12/06/2019	
					CT1685	Cardiac pacemakers	F1905	Device Revision or Replacement	A160601	Premature Elective Replacement Indicator	12/06/2019	
					CT257	Implantable cardiac pacemakers	F1905	Device Revision or Replacement	A160601	Premature Elective Replacement Indicator	12/06/2019	
					CT2866	Implantable cardiac electrical stimulators	F1905	Device Revision or Replacement	A160601	Premature Elective Replacement Indicator	12/06/2019	
9224	2020	All other reportable incidents	35852	Single-chamber implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	15/05/2020	15/05/2020
							F12	Serious Injury/ Illness/ Impairment	A070504	Premature Discharge of Battery	15/05/2020	15/05/2020
							F19	Surgical Intervention	A070504	Premature Discharge of Battery	15/05/2020	15/05/2020
9225	2020	All other reportable incidents	36241	Pacemaker/defibrillator lead	CT1684	Cardiac pacemakers and associated devices	E2104	Electric Shock	A070909	Oversensing		
									A071302	Inappropriate/Inadequate Shock/Stimulation		
							F08	Hospitalization or Prolonged Hospitalization	A070909	Oversensing		
									A071302	Inappropriate/Inadequate Shock/Stimulation		
							F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A070909	Oversensing		
									A071302	Inappropriate/Inadequate Shock/Stimulation		

							F19	Surgical Intervention	A070909	Oversensing		
									A071302	Inappropriate/Inadequate Shock/Stimulation		
9226	2020	All other reportable incidents	35852	Single-chamber implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A070909	Oversensing		
									A071302	Inappropriate/Inadequate Shock/Stimulation		
							F1905	Device Revision or Replacement	A070909	Oversensing		
									A071302	Inappropriate/Inadequate Shock/Stimulation		
9227	2020	All other reportable incidents	35852	Single-chamber implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E2104	Electric Shock	A070909	Oversensing		
									A071302	Inappropriate/Inadequate Shock/Stimulation		
							F08	Hospitalization or Prolonged Hospitalization	A070909	Oversensing		
									A071302	Inappropriate/Inadequate Shock/Stimulation		
							F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A070909	Oversensing		
									A071302	Inappropriate/Inadequate Shock/Stimulation		
							F19	Surgical Intervention	A070909	Oversensing		
									A071302	Inappropriate/Inadequate Shock/Stimulation		
9228	2020	All other reportable incidents	35852	Single-chamber implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A070909	Oversensing		
							F26	No Health Consequences or Impact	A070909	Oversensing		
9229	2020	All other reportable incidents	36241	Pacemaker/defibrillator lead	CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070909	Oversensing	23/04/2020	23/04/2020
							F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A070909	Oversensing	23/04/2020	23/04/2020
							F19	Surgical Intervention	A070909	Oversensing	23/04/2020	23/04/2020
							F1905	Device Revision or Replacement	A070909	Oversensing	23/04/2020	23/04/2020
9230	2020	All other reportable incidents	37265	Dual-chamber implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	F21	Unexpected Deterioration	A071302	Inappropriate/Inadequate Shock/Stimulation		
9231	2020	All other reportable incidents	36241	Pacemaker/defibrillator lead	CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A071302	Inappropriate/Inadequate Shock/Stimulation	23/04/2020	23/04/2020
							F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A071302	Inappropriate/Inadequate Shock/Stimulation	23/04/2020	23/04/2020
							F19	Surgical Intervention	A071302	Inappropriate/Inadequate Shock/Stimulation	23/04/2020	23/04/2020
							F1905	Device Revision or Replacement	A071302	Inappropriate/Inadequate Shock/Stimulation	23/04/2020	23/04/2020
9232	2020	All other reportable incidents	35852	Single-chamber implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A070909	Oversensing		
							F26	No Health Consequences or Impact	A070909	Oversensing		
9233	2020	All other reportable incidents	35223	Endocardial/interventricular septal pacing lead	CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A07	Electrical /Electronic Property Problem		
									A072202	Low Impedance		
							F26	No Health Consequences or Impact	A07	Electrical /Electronic Property Problem		
									A072202	Low Impedance		
9234	2020	All other reportable incidents	17882	Physiologic-monitoring defibrillation system	CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070912	Incorrect Interpretation of Signal		
							F26	No Health Consequences or Impact	A070912	Incorrect Interpretation of Signal		
9235	2020	Death	17882	Physiologic-monitoring defibrillation system	CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070908	Failure to Sense		
							F02	Death	A070908	Failure to Sense		
							F21	Unexpected Deterioration	A070908	Failure to Sense		
9236	2020	All other reportable incidents	17882	Physiologic-monitoring defibrillation system	CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A07	Electrical /Electronic Property Problem		
							F27	Problem Identified During Non-Clinical Procedure	A07	Electrical /Electronic Property Problem		
9237	2019	All other reportable incidents	36241	Pacemaker/defibrillator lead	CT1684	Cardiac pacemakers and associated devices	F1905	Device Revision or Replacement	A072001	Electromagnetic Interference	21/01/2019	
9238	2019	Death	35224	External pacemaker, epicardial pacing	CT1684	Cardiac pacemakers and associated devices			A12	Connection Problem		
					CT1685	Cardiac pacemakers			A12	Connection Problem		
9239	2019	All other reportable incidents	36241	Pacemaker/defibrillator lead	CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A0510	Retraction Problem	10/06/2019	
9240	2019	All other reportable incidents	35852	Single-chamber implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A072201	High Impedance		
9241	2019	All other reportable incidents	47265	Dual-chamber implantable pacemaker, rate-responsive	CT141	Electrodes/Leads and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070910	Undersensing		
					CT145	Cardiovascular devices	E2403	No Clinical Signs, Symptoms or Conditions	A070910	Undersensing		
					CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070910	Undersensing		
					CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A070910	Undersensing		
					CT233	Surgical	E2403	No Clinical Signs, Symptoms or Conditions	A070910	Undersensing		
					CT2406	Implantable	E2403	No Clinical Signs, Symptoms or Conditions	A070910	Undersensing		
					CT2407	Active implantable	E2403	No Clinical Signs, Symptoms or Conditions	A070910	Undersensing		
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A070910	Undersensing		
					CT2727	Active implant devices/ancillaries	E2403	No Clinical Signs, Symptoms or Conditions	A070910	Undersensing		
					CT2750	Implantable cardiac electrical stimulators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070910	Undersensing		
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A070910	Undersensing		
					CT2985	Active	E2403	No Clinical Signs, Symptoms or Conditions	A070910	Undersensing		
					CT3001	Implantable metal	E2403	No Clinical Signs, Symptoms or Conditions	A070910	Undersensing		
					CT3002	Implantable synthetic polymer	E2403	No Clinical Signs, Symptoms or Conditions	A070910	Undersensing		
					CT3006	Implantable inorganic material	E2403	No Clinical Signs, Symptoms or Conditions	A070910	Undersensing		
					CT314	Electrically powered	E2403	No Clinical Signs, Symptoms or Conditions	A070910	Undersensing		
					CT321	Long-term surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A070910	Undersensing		
					CT449	Generators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070910	Undersensing		
					CT981	Single use	E2403	No Clinical Signs, Symptoms or Conditions	A070910	Undersensing		
					CT983	Surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A070910	Undersensing		
9242	2019	All other reportable incidents	35852	Single-chamber implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A0722	Impedance Problem		
9243	2019	All other reportable incidents	17882	Physiologic-monitoring defibrillation system	CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A071301	Failure to Deliver Shock/Stimulation		
9244	2020	All other reportable incidents	36241	Pacemaker/defibrillator lead	CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A072201	High Impedance		
							F26	No Health Consequences or Impact	A072201	High Impedance		
9245	2020	All other reportable incidents	37265	Dual-chamber implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A0712	Pacing Problem		
									A071302	Inappropriate/Inadequate Shock/Stimulation		
							F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A0712	Pacing Problem		
									A071302	Inappropriate/Inadequate Shock/Stimulation		
							F14	Prolonged Episode of Care	A0712	Pacing Problem		
									A071302	Inappropriate/Inadequate Shock/Stimulation		
							F1905	Device Revision or Replacement	A0712	Pacing Problem		
									A071302	Inappropriate/Inadequate Shock/Stimulation		
							F26	No Health Consequences or Impact	A0712	Pacing Problem		
									A071302	Inappropriate/Inadequate Shock/Stimulation		
9246	2020	All other reportable incidents	37265	Dual-chamber implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E2104	Electric Shock	A0712	Pacing Problem		
									A071202	Inaccurate Synchronization		
									A071302	Inappropriate/Inadequate Shock/Stimulation		
							E2403	No Clinical Signs, Symptoms or Conditions	A0712	Pacing Problem		
									A071202	Inaccurate Synchronization		
									A071302	Inappropriate/Inadequate Shock/Stimulation		
							F26	No Health Consequences or Impact	A0712	Pacing Problem		
									A071202	Inaccurate Synchronization		
									A071302	Inappropriate/Inadequate Shock/Stimulation		
9247	2019	All other reportable incidents	47265	Dual-chamber implantable pacemaker, rate-	CT141	Electrodes/Leads and associated devices	F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery		

				responsive	CT145	Cardiovascular devices	F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery		
					CT1684	Cardiac pacemakers and associated devices	F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery		
					CT1685	Cardiac pacemakers	F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery		
					CT233	Surgical	F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery		
					CT2406	Implantable	F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery		
					CT2407	Active implantable	F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery		
					CT257	Implantable cardiac pacemakers	F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery		
					CT2727	Active implant devices/ancillaries	F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery		
					CT2750	Implantable cardiac electrical stimulators and associated devices	F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery		
					CT2866	Implantable cardiac electrical stimulators	F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery		
					CT2985	Active	F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery		
					CT3001	Implantable metal	F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery		
					CT3002	Implantable synthetic polymer	F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery		
					CT3006	Implantable inorganic material	F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery		
					CT314	Electrically powered	F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery		
					CT321	Long-term surgical invasive	F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery		
					CT449	Generators and associated devices	F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery		
					CT981	Single use	F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery		
					CT983	Surgical invasive	F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery		
9248	2019	All other reportable incidents	47265	Dual-chamber implantable pacemaker, rate-responsive	CT141	Electrodes/Leads and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
					CT145	Cardiovascular devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
					CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
					CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
					CT233	Surgical	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
					CT2406	Implantable	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
					CT2407	Active implantable	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
					CT2727	Active implant devices/ancillaries	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
					CT2750	Implantable cardiac electrical stimulators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
					CT2985	Active	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
					CT3001	Implantable metal	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
					CT3002	Implantable synthetic polymer	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
					CT3006	Implantable inorganic material	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
					CT314	Electrically powered	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
					CT321	Long-term surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
					CT449	Generators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
					CT981	Single use	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
					CT983	Surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
9249	2019	All other reportable incidents	47265	Dual-chamber implantable pacemaker, rate-responsive	CT141	Electrodes/Leads and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A072201	High Impedance		
					CT145	Cardiovascular devices	E2403	No Clinical Signs, Symptoms or Conditions	A072201	High Impedance		
					CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A072201	High Impedance		
					CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A072201	High Impedance		
					CT233	Surgical	E2403	No Clinical Signs, Symptoms or Conditions	A072201	High Impedance		
					CT2406	Implantable	E2403	No Clinical Signs, Symptoms or Conditions	A072201	High Impedance		
					CT2407	Active implantable	E2403	No Clinical Signs, Symptoms or Conditions	A072201	High Impedance		
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A072201	High Impedance		
					CT2727	Active implant devices/ancillaries	E2403	No Clinical Signs, Symptoms or Conditions	A072201	High Impedance		
					CT2750	Implantable cardiac electrical stimulators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A072201	High Impedance		
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A072201	High Impedance		
					CT2985	Active	E2403	No Clinical Signs, Symptoms or Conditions	A072201	High Impedance		
					CT3001	Implantable metal	E2403	No Clinical Signs, Symptoms or Conditions	A072201	High Impedance		
					CT3002	Implantable synthetic polymer	E2403	No Clinical Signs, Symptoms or Conditions	A072201	High Impedance		
					CT3006	Implantable inorganic material	E2403	No Clinical Signs, Symptoms or Conditions	A072201	High Impedance		
					CT314	Electrically powered	E2403	No Clinical Signs, Symptoms or Conditions	A072201	High Impedance		
					CT321	Long-term surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A072201	High Impedance		
					CT449	Generators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A072201	High Impedance		
					CT981	Single use	E2403	No Clinical Signs, Symptoms or Conditions	A072201	High Impedance		
					CT983	Surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A072201	High Impedance		
9250	2020	All other reportable incidents	37265	Dual-chamber implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E0601	Arrhythmia	A070909	Oversensing		
									A0712	Pacing Problem		
									A071202	Inaccurate Synchronization		
									A071302	Inappropriate/Inadequate Shock/Stimulation		
							E060109	Tachycardia	A070909	Oversensing		
									A0712	Pacing Problem		
									A071202	Inaccurate Synchronization		
									A071302	Inappropriate/Inadequate Shock/Stimulation		
							E2104	Electric Shock	A070909	Oversensing		
									A0712	Pacing Problem		
									A071202	Inaccurate Synchronization		
									A071302	Inappropriate/Inadequate Shock/Stimulation		
							F15	Recognized Device or Procedural Complication	A070909	Oversensing		
									A0712	Pacing Problem		
									A071202	Inaccurate Synchronization		
									A071302	Inappropriate/Inadequate Shock/Stimulation		
9251	2019	All other reportable incidents	47264	Dual-chamber implantable pacemaker, demand	CT141	Electrodes/Leads and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A0510	Retraction Problem	10/06/2019	
					CT145	Cardiovascular devices	E2403	No Clinical Signs, Symptoms or Conditions	A0510	Retraction Problem	10/06/2019	
					CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A0510	Retraction Problem	10/06/2019	
					CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A0510	Retraction Problem	10/06/2019	
					CT233	Surgical	E2403	No Clinical Signs, Symptoms or Conditions	A0510	Retraction Problem	10/06/2019	
					CT2406	Implantable	E2403	No Clinical Signs, Symptoms or Conditions	A0510	Retraction Problem	10/06/2019	
					CT2407	Active implantable	E2403	No Clinical Signs, Symptoms or Conditions	A0510	Retraction Problem	10/06/2019	
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A0510	Retraction Problem	10/06/2019	
					CT2727	Active implant devices/ancillaries	E2403	No Clinical Signs, Symptoms or Conditions	A0510	Retraction Problem	10/06/2019	
					CT2750	Implantable cardiac electrical stimulators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A0510	Retraction Problem	10/06/2019	

					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A0510	Retraction Problem	10/06/2019	
					CT2985	Active	E2403	No Clinical Signs, Symptoms or Conditions	A0510	Retraction Problem	10/06/2019	
					CT3001	Implantable metal	E2403	No Clinical Signs, Symptoms or Conditions	A0510	Retraction Problem	10/06/2019	
					CT3002	Implantable synthetic polymer	E2403	No Clinical Signs, Symptoms or Conditions	A0510	Retraction Problem	10/06/2019	
					CT3006	Implantable inorganic material	E2403	No Clinical Signs, Symptoms or Conditions	A0510	Retraction Problem	10/06/2019	
					CT314	Electrically powered	E2403	No Clinical Signs, Symptoms or Conditions	A0510	Retraction Problem	10/06/2019	
					CT321	Long-term surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A0510	Retraction Problem	10/06/2019	
					CT449	Generators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A0510	Retraction Problem	10/06/2019	
					CT981	Single use	E2403	No Clinical Signs, Symptoms or Conditions	A0510	Retraction Problem	10/06/2019	
					CT983	Surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A0510	Retraction Problem	10/06/2019	
9252	2020	All other reportable incidents	17882	Physiologic-monitoring defibrillation system	CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A0713	Defibrillation/Stimulation Problem		
							F27	Problem Identified During Non-Clinical Procedure	A0713	Defibrillation/Stimulation Problem		
9253	2019	All other reportable incidents	35852	Single-chamber implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A070909	Oversensing		
9254	2019	All other reportable incidents	35852	Single-chamber implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A071302	Inappropriate/Inadequate Shock/Stimulation		
9255	2019	All other reportable incidents	47265	Dual-chamber implantable pacemaker, rate-responsive	CT141	Electrodes/Leads and associated devices	F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A070909	Oversensing		
					CT145	Cardiovascular devices	F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A070909	Oversensing		
					CT1684	Cardiac pacemakers and associated devices	F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A070909	Oversensing		
					CT1685	Cardiac pacemakers	F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A070909	Oversensing		
					CT233	Surgical	F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A070909	Oversensing		
					CT2406	Implantable	F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A070909	Oversensing		
					CT2407	Active implantable	F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A070909	Oversensing		
					CT257	Implantable cardiac pacemakers	F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A070909	Oversensing		
					CT2727	Active implant devices/ancillaries	F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A070909	Oversensing		
					CT2750	Implantable cardiac electrical stimulators and associated devices	F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A070909	Oversensing		
					CT2866	Implantable cardiac electrical stimulators	F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A070909	Oversensing		
					CT2985	Active	F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A070909	Oversensing		
					CT3001	Implantable metal	F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A070909	Oversensing		
					CT3002	Implantable synthetic polymer	F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A070909	Oversensing		
					CT3006	Implantable inorganic material	F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A070909	Oversensing		
					CT314	Electrically powered	F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A070909	Oversensing		
					CT321	Long-term surgical invasive	F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A070909	Oversensing		
					CT449	Generators and associated devices	F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A070909	Oversensing		
					CT981	Single use	F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A070909	Oversensing		
					CT983	Surgical invasive	F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A070909	Oversensing		
9256	2019	All other reportable incidents	36241	Pacemaker/defibrillator lead	CT1684	Cardiac pacemakers and associated devices	F1905	Device Revision or Replacement	A070909	Oversensing	10/06/2019	
9257	2019	All other reportable incidents	47270	Cardiac resynchronization therapy implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	F1905	Device Revision or Replacement	A010102	Device Appears to Trigger Rejection	05/06/2019	
9258	2020	All other reportable incidents	35852	Single-chamber implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A070909	Oversensing		
9259	2020	All other reportable incidents	37265	Dual-chamber implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E2104	Electric Shock	A070910	Undersensing		
									A0712	Pacing Problem		
									A071302	Inappropriate/Inadequate Shock/Stimulation		
							F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A070910	Undersensing		
									A0712	Pacing Problem		
									A071302	Inappropriate/Inadequate Shock/Stimulation		
							F26	No Health Consequences or Impact	A070910	Undersensing		
									A0712	Pacing Problem		
									A071302	Inappropriate/Inadequate Shock/Stimulation		
9260	2020	All other reportable incidents	37265	Dual-chamber implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	14/05/2020	14/05/2020
							F1901	Additional Surgery	A070504	Premature Discharge of Battery	14/05/2020	14/05/2020
9261	2020	All other reportable incidents	36241	Pacemaker/defibrillator lead	CT1684	Cardiac pacemakers and associated devices	E2015	Unspecified Tissue Injury	A04	Material Integrity Problem	09/05/2020	09/05/2020
							E2114	Perforation	A04	Material Integrity Problem	09/05/2020	09/05/2020
							F12	Serious Injury/ Illness/ Impairment	A04	Material Integrity Problem	09/05/2020	09/05/2020
9262	2020	All other reportable incidents	35852	Single-chamber implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	12/05/2020	12/05/2020
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	12/05/2020	12/05/2020
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery	12/05/2020	12/05/2020
9263	2019	All other reportable incidents	37265	Dual-chamber implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A072201	High Impedance		
9264	2019	All other reportable incidents	36241	Pacemaker/defibrillator lead	CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A072201	High Impedance		
9265	2020	All other reportable incidents	37265	Dual-chamber implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	15/05/2020	15/05/2020
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	15/05/2020	15/05/2020
							F19	Surgical Intervention	A070504	Premature Discharge of Battery	15/05/2020	15/05/2020
9266	2020	All other reportable incidents	37265	Dual-chamber implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	30/06/2020	30/06/2020
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	30/06/2020	30/06/2020
							F19	Surgical Intervention	A070504	Premature Discharge of Battery	30/06/2020	30/06/2020
9267	2020	All other reportable incidents	35852	Single-chamber implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	13/07/2020	13/07/2020
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	13/07/2020	13/07/2020
							F19	Surgical Intervention	A070504	Premature Discharge of Battery	13/07/2020	13/07/2020
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	13/07/2020	13/07/2020
9268	2020	All other reportable incidents	37265	Dual-chamber implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	27/05/2020	27/05/2020
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	27/05/2020	27/05/2020
							F19	Surgical Intervention	A070504	Premature Discharge of Battery	27/05/2020	27/05/2020
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	27/05/2020	27/05/2020
9269	2019	All other reportable incidents	47265	Dual-chamber implantable pacemaker, rate-responsive	CT141	Electrodes/Leads and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
					CT145	Cardiovascular devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
					CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
					CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
					CT233	Surgical	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
					CT2406	Implantable	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
					CT2407	Active implantable	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
					CT2727	Active implant devices/ancillaries	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
					CT2750	Implantable cardiac electrical stimulators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
					CT2985	Active	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
					CT3001	Implantable metal	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
					CT3002	Implantable synthetic polymer	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
					CT3006	Implantable inorganic material	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		

					CT314	Electrically powered	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
					CT321	Long-term surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
					CT449	Generators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
					CT981	Single use	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
					CT983	Surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
9270	2019	All other reportable incidents	36241	Pacemaker/defibrillator lead	CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A010402	Migration	11/06/2019	
9271	2019	All other reportable incidents	47270	Cardiac resynchronization therapy implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A070909	Oversensing		
9272	2020	All other reportable incidents	36241	Pacemaker/defibrillator lead	CT1684	Cardiac pacemakers and associated devices	E0112	Dizziness	A24	No Apparent Device Problem		
							E2114	Perforation	A24	No Apparent Device Problem		
							E2330	Pain	A24	No Apparent Device Problem		
							F1905	Device Revision or Replacement	A24	No Apparent Device Problem		
9273	2019	All other reportable incidents	47270	Cardiac resynchronization therapy implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	02/07/2019	
9274	2020	All other reportable incidents	35852	Single-chamber implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E060109	Tachycardia	A0712	Pacing Problem		
									A071202	Inaccurate Synchronization		
							E2403	No Clinical Signs, Symptoms or Conditions	A0712	Pacing Problem		
									A071202	Inaccurate Synchronization		
							F15	Recognized Device or Procedural Complication	A0712	Pacing Problem		
									A071202	Inaccurate Synchronization		
9275	2020	All other reportable incidents	35852	Single-chamber implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A072201	High Impedance		
							F19	Surgical Intervention	A072201	High Impedance		
9276	2019	All other reportable incidents	47267	Single-chamber implantable pacemaker, rate-responsive	CT141	Electrodes/Leads and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
					CT145	Cardiovascular devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
					CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
					CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
					CT233	Surgical	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
					CT2406	Implantable	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
					CT2407	Active implantable	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
					CT2727	Active implant devices/ancillaries	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
					CT2750	Implantable cardiac electrical stimulators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
					CT2985	Active	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
					CT3001	Implantable metal	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
					CT3002	Implantable synthetic polymer	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
					CT3006	Implantable inorganic material	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
					CT314	Electrically powered	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
					CT321	Long-term surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
					CT449	Generators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
					CT981	Single use	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
					CT983	Surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
9277	2020	All other reportable incidents	35039	Epicardial pacing lead	CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070101	Failure to Capture		
									A0709	Device Sensing Problem		
									A072202	Low Impedance		
							F26	No Health Consequences or Impact	A070101	Failure to Capture		
									A0709	Device Sensing Problem		
									A072202	Low Impedance		
9278	2019	All other reportable incidents	47265	Dual-chamber implantable pacemaker, rate-responsive	CT141	Electrodes/Leads and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070909	Oversensing		
					CT145	Cardiovascular devices	E2403	No Clinical Signs, Symptoms or Conditions	A070909	Oversensing		
					CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070909	Oversensing		
					CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A070909	Oversensing		
					CT233	Surgical	E2403	No Clinical Signs, Symptoms or Conditions	A070909	Oversensing		
					CT2406	Implantable	E2403	No Clinical Signs, Symptoms or Conditions	A070909	Oversensing		
					CT2407	Active implantable	E2403	No Clinical Signs, Symptoms or Conditions	A070909	Oversensing		
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A070909	Oversensing		
					CT2727	Active implant devices/ancillaries	E2403	No Clinical Signs, Symptoms or Conditions	A070909	Oversensing		
					CT2750	Implantable cardiac electrical stimulators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070909	Oversensing		
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A070909	Oversensing		
					CT2985	Active	E2403	No Clinical Signs, Symptoms or Conditions	A070909	Oversensing		
					CT3001	Implantable metal	E2403	No Clinical Signs, Symptoms or Conditions	A070909	Oversensing		
					CT3002	Implantable synthetic polymer	E2403	No Clinical Signs, Symptoms or Conditions	A070909	Oversensing		
					CT3006	Implantable inorganic material	E2403	No Clinical Signs, Symptoms or Conditions	A070909	Oversensing		
					CT314	Electrically powered	E2403	No Clinical Signs, Symptoms or Conditions	A070909	Oversensing		
					CT321	Long-term surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A070909	Oversensing		
					CT449	Generators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070909	Oversensing		
					CT981	Single use	E2403	No Clinical Signs, Symptoms or Conditions	A070909	Oversensing		
					CT983	Surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A070909	Oversensing		
9279	2019	All other reportable incidents	47265	Dual-chamber implantable pacemaker, rate-responsive	CT141	Electrodes/Leads and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A11	Computer Software Problem		
					CT145	Cardiovascular devices	E2403	No Clinical Signs, Symptoms or Conditions	A11	Computer Software Problem		
					CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A11	Computer Software Problem		
					CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A11	Computer Software Problem		
					CT233	Surgical	E2403	No Clinical Signs, Symptoms or Conditions	A11	Computer Software Problem		
					CT2406	Implantable	E2403	No Clinical Signs, Symptoms or Conditions	A11	Computer Software Problem		
					CT2407	Active implantable	E2403	No Clinical Signs, Symptoms or Conditions	A11	Computer Software Problem		
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A11	Computer Software Problem		
					CT2727	Active implant devices/ancillaries	E2403	No Clinical Signs, Symptoms or Conditions	A11	Computer Software Problem		
					CT2750	Implantable cardiac electrical stimulators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A11	Computer Software Problem		
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A11	Computer Software Problem		
					CT2985	Active	E2403	No Clinical Signs, Symptoms or Conditions	A11	Computer Software Problem		
					CT3001	Implantable metal	E2403	No Clinical Signs, Symptoms or Conditions	A11	Computer Software Problem		
					CT3002	Implantable synthetic polymer	E2403	No Clinical Signs, Symptoms or Conditions	A11	Computer Software Problem		
					CT3006	Implantable inorganic material	E2403	No Clinical Signs, Symptoms or Conditions	A11	Computer Software Problem		
					CT314	Electrically powered	E2403	No Clinical Signs, Symptoms or Conditions	A11	Computer Software Problem		
					CT321	Long-term surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A11	Computer Software Problem		
					CT449	Generators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A11	Computer Software Problem		

					CT981	Single use	E2403	No Clinical Signs, Symptoms or Conditions	A11	Computer Software Problem			
					CT983	Surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A11	Computer Software Problem			
9280	2020	All other reportable incidents	17882	Physiologic-monitoring defibrillation system	CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A0719	Unexpected Shutdown			
							F26	No Health Consequences or Impact	A0719	Unexpected Shutdown			
9281	2019	Death	47270	Cardiac resynchronization therapy implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E2401	Insufficient Information	A24	No Apparent Device Problem	04/06/2019		
9282	2020	All other reportable incidents	36241	Pacemaker/defibrillator lead	CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A072201	High Impedance			
							F26	No Health Consequences or Impact	A072201	High Impedance			
9283	2020	All other reportable incidents	35852	Single-chamber implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A070909	Oversensing			
							F26	No Health Consequences or Impact	A070909	Oversensing			
9284	2020	All other reportable incidents	36241	Pacemaker/defibrillator lead	CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A0203	Defective Device			
									A070909	Oversensing			
							F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A0203	Defective Device			
									A070909	Oversensing			
							F23	Unexpected Medical Intervention	A0203	Defective Device			
									A070909	Oversensing			
9285	2020	All other reportable incidents	36241	Pacemaker/defibrillator lead	CT1684	Cardiac pacemakers and associated devices	E2104	Electric Shock	A070909	Oversensing	18/08/2020	18/08/2020	
									A071302	Inappropriate/Inadequate Shock/Stimulation	18/08/2020	18/08/2020	
							E2403	No Clinical Signs, Symptoms or Conditions	A070909	Oversensing	18/08/2020	18/08/2020	
									A071302	Inappropriate/Inadequate Shock/Stimulation	18/08/2020	18/08/2020	
							F08	Hospitalization or Prolonged Hospitalization	A070909	Oversensing	18/08/2020	18/08/2020	
									A071302	Inappropriate/Inadequate Shock/Stimulation	18/08/2020	18/08/2020	
							F12	Serious Injury/ Illness/ Impairment	A070909	Oversensing	18/08/2020	18/08/2020	
									A071302	Inappropriate/Inadequate Shock/Stimulation	18/08/2020	18/08/2020	
							F19	Surgical Intervention	A070909	Oversensing	18/08/2020	18/08/2020	
									A071302	Inappropriate/Inadequate Shock/Stimulation	18/08/2020	18/08/2020	
9286	2020	All other reportable incidents	47270	Cardiac resynchronization therapy implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	F21	Unexpected Deterioration	A071202	Inaccurate Synchronization			
9287	2020	All other reportable incidents	35223	Endocardial/interventricular septal pacing lead	CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070909	Oversensing			
									A0722	Impedance Problem			
									A072201	High Impedance			
							F26	No Health Consequences or Impact	A070909	Oversensing			
									A0722	Impedance Problem			
									A072201	High Impedance			
9288	2020	All other reportable incidents	17882	Physiologic-monitoring defibrillation system	CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070803	Failure to Power Up			
							F27	Problem Identified During Non-Clinical Procedure	A070803	Failure to Power Up			
9289	2019	All other reportable incidents	35223	Endocardial/interventricular septal pacing lead	CT1684	Cardiac pacemakers and associated devices	F1905	Device Revision or Replacement	A071204	No Pacing	18/06/2019		
9290	2019	All other reportable incidents	47265	Dual-chamber implantable pacemaker, rate-responsive	CT141	Electrodes/Leads and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A072201	High Impedance			
					CT145	Cardiovascular devices	E2403	No Clinical Signs, Symptoms or Conditions	A072201	High Impedance			
					CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A072201	High Impedance			
					CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A072201	High Impedance			
					CT233	Surgical	E2403	No Clinical Signs, Symptoms or Conditions	A072201	High Impedance			
					CT2406	Implantable	E2403	No Clinical Signs, Symptoms or Conditions	A072201	High Impedance			
					CT2407	Active implantable	E2403	No Clinical Signs, Symptoms or Conditions	A072201	High Impedance			
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A072201	High Impedance			
					CT2727	Active implant devices/ancillaries	E2403	No Clinical Signs, Symptoms or Conditions	A072201	High Impedance			
					CT2750	Implantable cardiac electrical stimulators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A072201	High Impedance			
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A072201	High Impedance			
					CT2985	Active	E2403	No Clinical Signs, Symptoms or Conditions	A072201	High Impedance			
					CT3001	Implantable metal	E2403	No Clinical Signs, Symptoms or Conditions	A072201	High Impedance			
					CT3002	Implantable synthetic polymer	E2403	No Clinical Signs, Symptoms or Conditions	A072201	High Impedance			
					CT3006	Implantable inorganic material	E2403	No Clinical Signs, Symptoms or Conditions	A072201	High Impedance			
					CT314	Electrically powered	E2403	No Clinical Signs, Symptoms or Conditions	A072201	High Impedance			
					CT321	Long-term surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A072201	High Impedance			
					CT449	Generators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A072201	High Impedance			
					CT981	Single use	E2403	No Clinical Signs, Symptoms or Conditions	A072201	High Impedance			
					CT983	Surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A072201	High Impedance			
9291	2019	All other reportable incidents	35852	Single-chamber implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A072202	Low Impedance			
9292	2019	All other reportable incidents	36241	Pacemaker/defibrillator lead	CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A072201	High Impedance			
9293	2019	All other reportable incidents	17882	Physiologic-monitoring defibrillation system	CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A090204	Image Display Error/Artifact			
9294	2019	All other reportable incidents	35852	Single-chamber implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	F1905	Device Revision or Replacement	A0203	Defective Device			
9295	2020	All other reportable incidents	17882	Physiologic-monitoring defibrillation system	CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070803	Failure to Power Up			
							F27	Problem Identified During Non-Clinical Procedure	A070803	Failure to Power Up			
9296	2020	All other reportable incidents	35852	Single-chamber implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery			
9297	2019	All other reportable incidents	37265	Dual-chamber implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A070603	Failure to Charge	28/06/2019	28/06/2019	
							F1905	Device Revision or Replacement	A070603	Failure to Charge	28/06/2019	28/06/2019	
							F26	No Health Consequences or Impact	A070603	Failure to Charge	28/06/2019	28/06/2019	
9298	2019	All other reportable incidents	47270	Cardiac resynchronization therapy implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	11/07/2019		
9299	2019	All other reportable incidents	35852	Single-chamber implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A070909	Oversensing			
9300	2019	All other reportable incidents	35852	Single-chamber implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A070909	Oversensing	17/06/2019		
9301	2019	All other reportable incidents	36241	Pacemaker/defibrillator lead	CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070909	Oversensing			
9302	2019	All other reportable incidents	36241	Pacemaker/defibrillator lead	CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A072201	High Impedance			
9303	2019	All other reportable incidents	35852	Single-chamber implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A072201	High Impedance			
9304	2019	All other reportable incidents	47270	Cardiac resynchronization therapy implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A070909	Oversensing			
9305	2019	All other reportable incidents	17882	Physiologic-monitoring defibrillation system	CT1684	Cardiac pacemakers and associated devices	F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A070801	Complete Loss of Power			
9306	2019	All other reportable incidents	17882	Physiologic-monitoring defibrillation system	CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070801	Complete Loss of Power			
9307	2019	Death	47265	Dual-chamber implantable pacemaker, rate-responsive	CT141	Electrodes/Leads and associated devices	E2401	Insufficient Information	A24	No Apparent Device Problem			
					CT145	Cardiovascular devices	E2401	Insufficient Information	A24	No Apparent Device Problem			
					CT1684	Cardiac pacemakers and associated devices	E2401	Insufficient Information	A24	No Apparent Device Problem			
					CT1685	Cardiac pacemakers	E2401	Insufficient Information	A24	No Apparent Device Problem			
					CT233	Surgical	E2401	Insufficient Information	A24	No Apparent Device Problem			
					CT2406	Implantable	E2401	Insufficient Information	A24	No Apparent Device Problem			
					CT2407	Active implantable	E2401	Insufficient Information	A24	No Apparent Device Problem			
					CT257	Implantable cardiac pacemakers	E2401	Insufficient Information	A24	No Apparent Device Problem			
					CT2727	Active implant devices/ancillaries	E2401	Insufficient Information	A24	No Apparent Device Problem			

					CT2750	Implantable cardiac electrical stimulators and associated devices	E2401	Insufficient Information	A24	No Apparent Device Problem		
					CT2866	Implantable cardiac electrical stimulators	E2401	Insufficient Information	A24	No Apparent Device Problem		
					CT2985	Active	E2401	Insufficient Information	A24	No Apparent Device Problem		
					CT3001	Implantable metal	E2401	Insufficient Information	A24	No Apparent Device Problem		
					CT3002	Implantable synthetic polymer	E2401	Insufficient Information	A24	No Apparent Device Problem		
					CT3006	Implantable inorganic material	E2401	Insufficient Information	A24	No Apparent Device Problem		
					CT314	Electrically powered	E2401	Insufficient Information	A24	No Apparent Device Problem		
					CT321	Long-term surgical invasive	E2401	Insufficient Information	A24	No Apparent Device Problem		
					CT449	Generators and associated devices	E2401	Insufficient Information	A24	No Apparent Device Problem		
					CT981	Single use	E2401	Insufficient Information	A24	No Apparent Device Problem		
					CT983	Surgical invasive	E2401	Insufficient Information	A24	No Apparent Device Problem		
9308	2019	All other reportable incidents	35852	Single-chamber implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	27/06/2019	31/12/2019
9309	2019	All other reportable incidents	47265	Dual-chamber implantable pacemaker, rate-responsive	CT141	Electrodes/Leads and associated devices	F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	25/04/2019	
					CT145	Cardiovascular devices	F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	25/04/2019	
					CT1684	Cardiac pacemakers and associated devices	F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	25/04/2019	
					CT1685	Cardiac pacemakers	F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	25/04/2019	
					CT233	Surgical	F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	25/04/2019	
					CT2406	Implantable	F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	25/04/2019	
					CT2407	Active implantable	F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	25/04/2019	
					CT257	Implantable cardiac pacemakers	F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	25/04/2019	
					CT2727	Active implant devices/ancillaries	F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	25/04/2019	
					CT2750	Implantable cardiac electrical stimulators and associated devices	F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	25/04/2019	
					CT2866	Implantable cardiac electrical stimulators	F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	25/04/2019	
					CT2985	Active	F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	25/04/2019	
					CT3001	Implantable metal	F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	25/04/2019	
					CT3002	Implantable synthetic polymer	F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	25/04/2019	
					CT3006	Implantable inorganic material	F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	25/04/2019	
					CT314	Electrically powered	F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	25/04/2019	
					CT321	Long-term surgical invasive	F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	25/04/2019	
					CT449	Generators and associated devices	F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	25/04/2019	
					CT981	Single use	F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	25/04/2019	
					CT983	Surgical invasive	F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	25/04/2019	
9310	2019	All other reportable incidents	47267	Single-chamber implantable pacemaker, rate-responsive	CT141	Electrodes/Leads and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	15/07/2019	
					CT145	Cardiovascular devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	15/07/2019	
					CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	15/07/2019	
					CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	15/07/2019	
					CT233	Surgical	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	15/07/2019	
					CT2406	Implantable	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	15/07/2019	
					CT2407	Active implantable	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	15/07/2019	
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	15/07/2019	
					CT2727	Active implant devices/ancillaries	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	15/07/2019	
					CT2750	Implantable cardiac electrical stimulators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	15/07/2019	
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	15/07/2019	
					CT2985	Active	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	15/07/2019	
					CT3001	Implantable metal	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	15/07/2019	
					CT3002	Implantable synthetic polymer	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	15/07/2019	
					CT3006	Implantable inorganic material	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	15/07/2019	
					CT314	Electrically powered	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	15/07/2019	
					CT321	Long-term surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	15/07/2019	
					CT449	Generators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	15/07/2019	
					CT981	Single use	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	15/07/2019	
					CT983	Surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	15/07/2019	
9311	2019	All other reportable incidents	35852	Single-chamber implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A071302	Inappropriate/Inadequate Shock/Stimulation		
9312	2019	All other reportable incidents	47270	Cardiac resynchronization therapy implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A070909	Oversensing		
9313	2019	All other reportable incidents	47265	Dual-chamber implantable pacemaker, rate-responsive	CT141	Electrodes/Leads and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
					CT145	Cardiovascular devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
					CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
					CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
					CT233	Surgical	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
					CT2406	Implantable	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
					CT2407	Active implantable	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
					CT2727	Active implant devices/ancillaries	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
					CT2750	Implantable cardiac electrical stimulators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
					CT2985	Active	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
					CT3001	Implantable metal	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
					CT3002	Implantable synthetic polymer	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
					CT3006	Implantable inorganic material	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
					CT314	Electrically powered	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
					CT321	Long-term surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
					CT449	Generators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
					CT981	Single use	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
					CT983	Surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
9314	2019	All other reportable incidents	37265	Dual-chamber implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	F1905	Device Revision or Replacement	A0512	Unintended Movement	13/08/2019	
9315	2019	All other reportable incidents	36241	Pacemaker/defibrillator lead	CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A0712	Pacing Problem		
9316	2020	All other reportable incidents	47270	Cardiac resynchronization therapy implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	11/06/2020	11/06/2020
							F1901	Additional Surgery	A070504	Premature Discharge of Battery	11/06/2020	11/06/2020
9317	2020	All other reportable incidents	47270	Cardiac resynchronization therapy implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A27	Appropriate Device Problem Term/Code not Available		
							F26	No Health Consequences or Impact	A27	Appropriate Device Problem Term/Code not Available		
9318	2020	All other reportable incidents	47265	Dual-chamber implantable pacemaker, rate-responsive	CT141	Electrodes/Leads and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A071102	Failure to Interrogate	22/05/2020	22/05/2020
							F1901	Additional Surgery	A071102	Failure to Interrogate	22/05/2020	22/05/2020

					CT145	Cardiovascular devices	E2403	No Clinical Signs, Symptoms or Conditions	A071102	Failure to Interrogate	22/05/2020	22/05/2020
							F1901	Additional Surgery	A071102	Failure to Interrogate	22/05/2020	22/05/2020
					CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A071102	Failure to Interrogate	22/05/2020	22/05/2020
							F1901	Additional Surgery	A071102	Failure to Interrogate	22/05/2020	22/05/2020
					CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A071102	Failure to Interrogate	22/05/2020	22/05/2020
							F1901	Additional Surgery	A071102	Failure to Interrogate	22/05/2020	22/05/2020
					CT233	Surgical	E2403	No Clinical Signs, Symptoms or Conditions	A071102	Failure to Interrogate	22/05/2020	22/05/2020
							F1901	Additional Surgery	A071102	Failure to Interrogate	22/05/2020	22/05/2020
					CT2406	Implantable	E2403	No Clinical Signs, Symptoms or Conditions	A071102	Failure to Interrogate	22/05/2020	22/05/2020
							F1901	Additional Surgery	A071102	Failure to Interrogate	22/05/2020	22/05/2020
					CT2407	Active implantable	E2403	No Clinical Signs, Symptoms or Conditions	A071102	Failure to Interrogate	22/05/2020	22/05/2020
							F1901	Additional Surgery	A071102	Failure to Interrogate	22/05/2020	22/05/2020
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A071102	Failure to Interrogate	22/05/2020	22/05/2020
							F1901	Additional Surgery	A071102	Failure to Interrogate	22/05/2020	22/05/2020
					CT2727	Active implant devices/ancillaries	E2403	No Clinical Signs, Symptoms or Conditions	A071102	Failure to Interrogate	22/05/2020	22/05/2020
							F1901	Additional Surgery	A071102	Failure to Interrogate	22/05/2020	22/05/2020
					CT2750	Implantable cardiac electrical stimulators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A071102	Failure to Interrogate	22/05/2020	22/05/2020
							F1901	Additional Surgery	A071102	Failure to Interrogate	22/05/2020	22/05/2020
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A071102	Failure to Interrogate	22/05/2020	22/05/2020
							F1901	Additional Surgery	A071102	Failure to Interrogate	22/05/2020	22/05/2020
					CT2985	Active	E2403	No Clinical Signs, Symptoms or Conditions	A071102	Failure to Interrogate	22/05/2020	22/05/2020
							F1901	Additional Surgery	A071102	Failure to Interrogate	22/05/2020	22/05/2020
					CT3001	Implantable metal	E2403	No Clinical Signs, Symptoms or Conditions	A071102	Failure to Interrogate	22/05/2020	22/05/2020
							F1901	Additional Surgery	A071102	Failure to Interrogate	22/05/2020	22/05/2020
					CT3002	Implantable synthetic polymer	E2403	No Clinical Signs, Symptoms or Conditions	A071102	Failure to Interrogate	22/05/2020	22/05/2020
							F1901	Additional Surgery	A071102	Failure to Interrogate	22/05/2020	22/05/2020
					CT3006	Implantable inorganic material	E2403	No Clinical Signs, Symptoms or Conditions	A071102	Failure to Interrogate	22/05/2020	22/05/2020
							F1901	Additional Surgery	A071102	Failure to Interrogate	22/05/2020	22/05/2020
					CT314	Electrically powered	E2403	No Clinical Signs, Symptoms or Conditions	A071102	Failure to Interrogate	22/05/2020	22/05/2020
							F1901	Additional Surgery	A071102	Failure to Interrogate	22/05/2020	22/05/2020
					CT321	Long-term surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A071102	Failure to Interrogate	22/05/2020	22/05/2020
							F1901	Additional Surgery	A071102	Failure to Interrogate	22/05/2020	22/05/2020
					CT449	Generators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A071102	Failure to Interrogate	22/05/2020	22/05/2020
							F1901	Additional Surgery	A071102	Failure to Interrogate	22/05/2020	22/05/2020
					CT981	Single use	E2403	No Clinical Signs, Symptoms or Conditions	A071102	Failure to Interrogate	22/05/2020	22/05/2020
							F1901	Additional Surgery	A071102	Failure to Interrogate	22/05/2020	22/05/2020
					CT983	Surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A071102	Failure to Interrogate	22/05/2020	22/05/2020
							F1901	Additional Surgery	A071102	Failure to Interrogate	22/05/2020	22/05/2020
9319	2020	All other reportable incidents	47265	Dual-chamber implantable pacemaker, rate-responsive	CT141	Electrodes/Leads and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A0712	Pacing Problem		
							F26	No Health Consequences or Impact	A0712	Pacing Problem		
					CT145	Cardiovascular devices	E2403	No Clinical Signs, Symptoms or Conditions	A0712	Pacing Problem		
							F26	No Health Consequences or Impact	A0712	Pacing Problem		
					CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A0712	Pacing Problem		
							F26	No Health Consequences or Impact	A0712	Pacing Problem		
					CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A0712	Pacing Problem		
							F26	No Health Consequences or Impact	A0712	Pacing Problem		
					CT233	Surgical	E2403	No Clinical Signs, Symptoms or Conditions	A0712	Pacing Problem		
							F26	No Health Consequences or Impact	A0712	Pacing Problem		
					CT2406	Implantable	E2403	No Clinical Signs, Symptoms or Conditions	A0712	Pacing Problem		
							F26	No Health Consequences or Impact	A0712	Pacing Problem		
					CT2407	Active implantable	E2403	No Clinical Signs, Symptoms or Conditions	A0712	Pacing Problem		
							F26	No Health Consequences or Impact	A0712	Pacing Problem		
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A0712	Pacing Problem		
							F26	No Health Consequences or Impact	A0712	Pacing Problem		
					CT2727	Active implant devices/ancillaries	E2403	No Clinical Signs, Symptoms or Conditions	A0712	Pacing Problem		
							F26	No Health Consequences or Impact	A0712	Pacing Problem		
					CT2750	Implantable cardiac electrical stimulators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A0712	Pacing Problem		
							F26	No Health Consequences or Impact	A0712	Pacing Problem		
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A0712	Pacing Problem		
							F26	No Health Consequences or Impact	A0712	Pacing Problem		
					CT2985	Active	E2403	No Clinical Signs, Symptoms or Conditions	A0712	Pacing Problem		
							F26	No Health Consequences or Impact	A0712	Pacing Problem		
					CT3001	Implantable metal	E2403	No Clinical Signs, Symptoms or Conditions	A0712	Pacing Problem		
							F26	No Health Consequences or Impact	A0712	Pacing Problem		
					CT3002	Implantable synthetic polymer	E2403	No Clinical Signs, Symptoms or Conditions	A0712	Pacing Problem		
							F26	No Health Consequences or Impact	A0712	Pacing Problem		
					CT3006	Implantable inorganic material	E2403	No Clinical Signs, Symptoms or Conditions	A0712	Pacing Problem		
							F26	No Health Consequences or Impact	A0712	Pacing Problem		
					CT314	Electrically powered	E2403	No Clinical Signs, Symptoms or Conditions	A0712	Pacing Problem		
							F26	No Health Consequences or Impact	A0712	Pacing Problem		
					CT321	Long-term surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A0712	Pacing Problem		
							F26	No Health Consequences or Impact	A0712	Pacing Problem		
					CT449	Generators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A0712	Pacing Problem		
							F26	No Health Consequences or Impact	A0712	Pacing Problem		
					CT981	Single use	E2403	No Clinical Signs, Symptoms or Conditions	A0712	Pacing Problem		
							F26	No Health Consequences or Impact	A0712	Pacing Problem		
					CT983	Surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A0712	Pacing Problem		
							F26	No Health Consequences or Impact	A0712	Pacing Problem		
9320	2020	All other reportable incidents	35852	Single-chamber implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A071302	Inappropriate/Inadequate Shock/Stimulation		
							F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A071302	Inappropriate/Inadequate Shock/Stimulation		
							F26	No Health Consequences or Impact	A071302	Inappropriate/Inadequate Shock/Stimulation		
9321	2020	All other reportable incidents	47270	Cardiac resynchronization therapy implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A070909	Oversensing		
							F26	No Health Consequences or Impact	A070909	Oversensing		
9322	2019	All other reportable incidents	37265	Dual-chamber implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A070909	Oversensing		
9323	2020	All other reportable incidents	47263	Cardiac resynchronization therapy implantable pacemaker	CT1684	Cardiac pacemakers and associated devices	E0112	Dizziness	A24	No Apparent Device Problem		
							E2007	Fall	A24	No Apparent Device Problem		
							E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem		
							F26	No Health Consequences or Impact	A24	No Apparent Device Problem		

					CT1685	Cardiac pacemakers	E0112	Dizziness	A24	No Apparent Device Problem		
							E2007	Fall	A24	No Apparent Device Problem		
							E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem		
							F26	No Health Consequences or Impact	A24	No Apparent Device Problem		
					CT257	Implantable cardiac pacemakers	E0112	Dizziness	A24	No Apparent Device Problem		
							E2007	Fall	A24	No Apparent Device Problem		
							E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem		
							F26	No Health Consequences or Impact	A24	No Apparent Device Problem		
					CT2866	Implantable cardiac electrical stimulators	E0112	Dizziness	A24	No Apparent Device Problem		
							E2007	Fall	A24	No Apparent Device Problem		
							E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem		
							F26	No Health Consequences or Impact	A24	No Apparent Device Problem		
9324	2020	All other reportable incidents	35852	Single-chamber implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A070909	Oversensing		
							F26	No Health Consequences or Impact	A070909	Oversensing		
9325	2020	All other reportable incidents	47265	Dual-chamber implantable pacemaker, rate-responsive	CT141	Electrodes/Leads and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
									A090804	False Positive Result		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
									A090804	False Positive Result		
					CT145	Cardiovascular devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
									A090804	False Positive Result		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
									A090804	False Positive Result		
					CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
									A090804	False Positive Result		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
									A090804	False Positive Result		
					CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
									A090804	False Positive Result		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
									A090804	False Positive Result		
					CT233	Surgical	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
									A090804	False Positive Result		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
									A090804	False Positive Result		
					CT2406	Implantable	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
									A090804	False Positive Result		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
									A090804	False Positive Result		
					CT2407	Active implantable	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
									A090804	False Positive Result		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
									A090804	False Positive Result		
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
									A090804	False Positive Result		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
									A090804	False Positive Result		
					CT2727	Active implant devices/ancillaries	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
									A090804	False Positive Result		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
									A090804	False Positive Result		
					CT2750	Implantable cardiac electrical stimulators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
									A090804	False Positive Result		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
									A090804	False Positive Result		
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
									A090804	False Positive Result		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
									A090804	False Positive Result		
					CT2985	Active	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
									A090804	False Positive Result		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
									A090804	False Positive Result		
					CT3001	Implantable metal	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
									A090804	False Positive Result		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
									A090804	False Positive Result		
					CT3002	Implantable synthetic polymer	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
									A090804	False Positive Result		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
									A090804	False Positive Result		
					CT3006	Implantable inorganic material	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
									A090804	False Positive Result		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
									A090804	False Positive Result		
					CT314	Electrically powered	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
									A090804	False Positive Result		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
									A090804	False Positive Result		
					CT321	Long-term surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
									A090804	False Positive Result		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
									A090804	False Positive Result		
					CT449	Generators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
									A090804	False Positive Result		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
									A090804	False Positive Result		
					CT981	Single use	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
									A090804	False Positive Result		

							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
									A090804	False Positive Result		
					CT983	Surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
									A090804	False Positive Result		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
									A090804	False Positive Result		
9326	2020	All other reportable incidents	35852	Single-chamber implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A070909	Oversensing		
							F14	Prolonged Episode of Care	A070909	Oversensing		
							F23	Unexpected Medical Intervention	A070909	Oversensing		
9327	2019	All other reportable incidents	47270	Cardiac resynchronization therapy implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A071302	Inappropriate/Inadequate Shock/Stimulation		
9328	2019	All other reportable incidents	36241	Pacemaker/defibrillator lead	CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A0510	Retraction Problem	21/06/2019	
9329	2020	All other reportable incidents	35852	Single-chamber implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	08/06/2020	08/06/2020
							F12	Serious Injury/ Illness/ Impairment	A070504	Premature Discharge of Battery	08/06/2020	08/06/2020
							F19	Surgical Intervention	A070504	Premature Discharge of Battery	08/06/2020	08/06/2020
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	08/06/2020	08/06/2020
9330	2020	All other reportable incidents	36241	Pacemaker/defibrillator lead	CT1684	Cardiac pacemakers and associated devices	E2104	Electric Shock	A071302	Inappropriate/Inadequate Shock/Stimulation	19/05/2020	19/05/2020
							F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A071302	Inappropriate/Inadequate Shock/Stimulation	19/05/2020	19/05/2020
							F1905	Device Revision or Replacement	A071302	Inappropriate/Inadequate Shock/Stimulation	19/05/2020	19/05/2020
9331	2020	All other reportable incidents	35852	Single-chamber implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E2104	Electric Shock	A071302	Inappropriate/Inadequate Shock/Stimulation		
							E2403	No Clinical Signs, Symptoms or Conditions	A071302	Inappropriate/Inadequate Shock/Stimulation		
							F26	No Health Consequences or Impact	A071302	Inappropriate/Inadequate Shock/Stimulation		
9332	2020	All other reportable incidents	35223	Endocardial/interventricular septal pacing lead	CT1684	Cardiac pacemakers and associated devices	E0717	Dyspnea	A051201	Device Dislodged or Dislocated		
									A070102	High Capture Threshold		
									A070103	Intermittent Capture		
									A070910	Undersensing		
									A0722	Impedance Problem		
									A1502	Positioning Problem		
							E0752	Unspecified Respiratory Problem	A051201	Device Dislodged or Dislocated		
									A070102	High Capture Threshold		
									A070103	Intermittent Capture		
									A070910	Undersensing		
									A0722	Impedance Problem		
									A1502	Positioning Problem		
							F1904	Device Repositioning	A051201	Device Dislodged or Dislocated		
									A070102	High Capture Threshold		
									A070103	Intermittent Capture		
									A070910	Undersensing		
									A0722	Impedance Problem		
									A1502	Positioning Problem		
							F1905	Device Revision or Replacement	A051201	Device Dislodged or Dislocated		
									A070102	High Capture Threshold		
									A070103	Intermittent Capture		
									A070910	Undersensing		
									A0722	Impedance Problem		
									A1502	Positioning Problem		
9333	2019	All other reportable incidents	35852	Single-chamber implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A070909	Oversensing	09/06/2020	09/06/2020
							F19	Surgical Intervention	A070909	Oversensing	09/06/2020	09/06/2020
9334	2019	All other reportable incidents	35852	Single-chamber implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E2401	Insufficient Information	A	Annex A		
9335	2019	All other reportable incidents	36241	Pacemaker/defibrillator lead	CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070909	Oversensing		
9336	2019	All other reportable incidents	35852	Single-chamber implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A072201	High Impedance		
9337	2019	All other reportable incidents	35852	Single-chamber implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E2401	Insufficient Information	A071201	Failure to Convert Rhythm	22/02/2021	22/02/2021
									A13	Communication or Transmission Problem	22/02/2021	22/02/2021
							E2403	No Clinical Signs, Symptoms or Conditions	A071201	Failure to Convert Rhythm	22/02/2021	22/02/2021
									A13	Communication or Transmission Problem	22/02/2021	22/02/2021
							F15	Recognized Device or Procedural Complication	A071201	Failure to Convert Rhythm	22/02/2021	22/02/2021
									A13	Communication or Transmission Problem	22/02/2021	22/02/2021
9338	2019	All other reportable incidents	35852	Single-chamber implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A071302	Inappropriate/Inadequate Shock/Stimulation		
9339	2019	All other reportable incidents	36241	Pacemaker/defibrillator lead	CT1684	Cardiac pacemakers and associated devices	F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A071302	Inappropriate/Inadequate Shock/Stimulation		
9340	2019	All other reportable incidents	36241	Pacemaker/defibrillator lead	CT1684	Cardiac pacemakers and associated devices	F1905	Device Revision or Replacement	A072201	High Impedance		
9341	2020	All other reportable incidents	35223	Endocardial/interventricular septal pacing lead	CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A040507	Naturally Worn	01/06/2020	01/06/2020
									A072201	High Impedance	01/06/2020	01/06/2020
							F1901	Additional Surgery	A040507	Naturally Worn	01/06/2020	01/06/2020
									A072201	High Impedance	01/06/2020	01/06/2020
							F1905	Device Revision or Replacement	A040507	Naturally Worn	01/06/2020	01/06/2020
									A072201	High Impedance	01/06/2020	01/06/2020
9342	2020	All other reportable incidents	47265	Dual-chamber implantable pacemaker, rate-responsive	CT141	Electrodes/Leads and associated devices	E0506	Hemorrhage/Blood Loss/Bleeding	A071102	Failure to Interrogate	01/06/2020	01/06/2020
							E2402	Appropriate Clinical Signs, Symptoms and Conditions Term/Code not Available	A071102	Failure to Interrogate	01/06/2020	01/06/2020
							F1905	Device Revision or Replacement	A071102	Failure to Interrogate	01/06/2020	01/06/2020
							F28	Appropriate Health Impact Term/Code not Available	A071102	Failure to Interrogate	01/06/2020	01/06/2020
					CT145	Cardiovascular devices	E0506	Hemorrhage/Blood Loss/Bleeding	A071102	Failure to Interrogate	01/06/2020	01/06/2020
							E2402	Appropriate Clinical Signs, Symptoms and Conditions Term/Code not Available	A071102	Failure to Interrogate	01/06/2020	01/06/2020
							F1905	Device Revision or Replacement	A071102	Failure to Interrogate	01/06/2020	01/06/2020
							F28	Appropriate Health Impact Term/Code not Available	A071102	Failure to Interrogate	01/06/2020	01/06/2020
					CT1684	Cardiac pacemakers and associated devices	E0506	Hemorrhage/Blood Loss/Bleeding	A071102	Failure to Interrogate	01/06/2020	01/06/2020
							E2402	Appropriate Clinical Signs, Symptoms and Conditions Term/Code not Available	A071102	Failure to Interrogate	01/06/2020	01/06/2020
							F1905	Device Revision or Replacement	A071102	Failure to Interrogate	01/06/2020	01/06/2020
							F28	Appropriate Health Impact Term/Code not Available	A071102	Failure to Interrogate	01/06/2020	01/06/2020
					CT1685	Cardiac pacemakers	E0506	Hemorrhage/Blood Loss/Bleeding	A071102	Failure to Interrogate	01/06/2020	01/06/2020
							E2402	Appropriate Clinical Signs, Symptoms and Conditions Term/Code not Available	A071102	Failure to Interrogate	01/06/2020	01/06/2020
							F1905	Device Revision or Replacement	A071102	Failure to Interrogate	01/06/2020	01/06/2020
							F28	Appropriate Health Impact Term/Code not Available	A071102	Failure to Interrogate	01/06/2020	01/06/2020
					CT233	Surgical	E0506	Hemorrhage/Blood Loss/Bleeding	A071102	Failure to Interrogate	01/06/2020	01/06/2020
							E2402	Appropriate Clinical Signs, Symptoms and Conditions Term/Code not Available	A071102	Failure to Interrogate	01/06/2020	01/06/2020

							F1905	Device Revision or Replacement	A071102	Failure to Interrogate	01/06/2020	01/06/2020
							F28	Appropriate Health Impact Term/Code not Available	A071102	Failure to Interrogate	01/06/2020	01/06/2020
					CT2406	Implantable	E0506	Hemorrhage/Blood Loss/Bleeding	A071102	Failure to Interrogate	01/06/2020	01/06/2020
							E2402	Appropriate Clinical Signs, Symptoms and Conditions Term/Code not Available	A071102	Failure to Interrogate	01/06/2020	01/06/2020
							F1905	Device Revision or Replacement	A071102	Failure to Interrogate	01/06/2020	01/06/2020
							F28	Appropriate Health Impact Term/Code not Available	A071102	Failure to Interrogate	01/06/2020	01/06/2020
					CT2407	Active implantable	E0506	Hemorrhage/Blood Loss/Bleeding	A071102	Failure to Interrogate	01/06/2020	01/06/2020
							E2402	Appropriate Clinical Signs, Symptoms and Conditions Term/Code not Available	A071102	Failure to Interrogate	01/06/2020	01/06/2020
							F1905	Device Revision or Replacement	A071102	Failure to Interrogate	01/06/2020	01/06/2020
							F28	Appropriate Health Impact Term/Code not Available	A071102	Failure to Interrogate	01/06/2020	01/06/2020
					CT257	Implantable cardiac pacemakers	E0506	Hemorrhage/Blood Loss/Bleeding	A071102	Failure to Interrogate	01/06/2020	01/06/2020
							E2402	Appropriate Clinical Signs, Symptoms and Conditions Term/Code not Available	A071102	Failure to Interrogate	01/06/2020	01/06/2020
							F1905	Device Revision or Replacement	A071102	Failure to Interrogate	01/06/2020	01/06/2020
							F28	Appropriate Health Impact Term/Code not Available	A071102	Failure to Interrogate	01/06/2020	01/06/2020
					CT2727	Active implant devices/ancillaries	E0506	Hemorrhage/Blood Loss/Bleeding	A071102	Failure to Interrogate	01/06/2020	01/06/2020
							E2402	Appropriate Clinical Signs, Symptoms and Conditions Term/Code not Available	A071102	Failure to Interrogate	01/06/2020	01/06/2020
							F1905	Device Revision or Replacement	A071102	Failure to Interrogate	01/06/2020	01/06/2020
							F28	Appropriate Health Impact Term/Code not Available	A071102	Failure to Interrogate	01/06/2020	01/06/2020
					CT2750	Implantable cardiac electrical stimulators and associated devices	E0506	Hemorrhage/Blood Loss/Bleeding	A071102	Failure to Interrogate	01/06/2020	01/06/2020
							E2402	Appropriate Clinical Signs, Symptoms and Conditions Term/Code not Available	A071102	Failure to Interrogate	01/06/2020	01/06/2020
							F1905	Device Revision or Replacement	A071102	Failure to Interrogate	01/06/2020	01/06/2020
							F28	Appropriate Health Impact Term/Code not Available	A071102	Failure to Interrogate	01/06/2020	01/06/2020
					CT2866	Implantable cardiac electrical stimulators	E0506	Hemorrhage/Blood Loss/Bleeding	A071102	Failure to Interrogate	01/06/2020	01/06/2020
							E2402	Appropriate Clinical Signs, Symptoms and Conditions Term/Code not Available	A071102	Failure to Interrogate	01/06/2020	01/06/2020
							F1905	Device Revision or Replacement	A071102	Failure to Interrogate	01/06/2020	01/06/2020
							F28	Appropriate Health Impact Term/Code not Available	A071102	Failure to Interrogate	01/06/2020	01/06/2020
					CT2985	Active	E0506	Hemorrhage/Blood Loss/Bleeding	A071102	Failure to Interrogate	01/06/2020	01/06/2020
							E2402	Appropriate Clinical Signs, Symptoms and Conditions Term/Code not Available	A071102	Failure to Interrogate	01/06/2020	01/06/2020
							F1905	Device Revision or Replacement	A071102	Failure to Interrogate	01/06/2020	01/06/2020
							F28	Appropriate Health Impact Term/Code not Available	A071102	Failure to Interrogate	01/06/2020	01/06/2020
					CT3001	Implantable metal	E0506	Hemorrhage/Blood Loss/Bleeding	A071102	Failure to Interrogate	01/06/2020	01/06/2020
							E2402	Appropriate Clinical Signs, Symptoms and Conditions Term/Code not Available	A071102	Failure to Interrogate	01/06/2020	01/06/2020
							F1905	Device Revision or Replacement	A071102	Failure to Interrogate	01/06/2020	01/06/2020
							F28	Appropriate Health Impact Term/Code not Available	A071102	Failure to Interrogate	01/06/2020	01/06/2020
					CT3002	Implantable synthetic polymer	E0506	Hemorrhage/Blood Loss/Bleeding	A071102	Failure to Interrogate	01/06/2020	01/06/2020
							E2402	Appropriate Clinical Signs, Symptoms and Conditions Term/Code not Available	A071102	Failure to Interrogate	01/06/2020	01/06/2020
							F1905	Device Revision or Replacement	A071102	Failure to Interrogate	01/06/2020	01/06/2020
							F28	Appropriate Health Impact Term/Code not Available	A071102	Failure to Interrogate	01/06/2020	01/06/2020
					CT3006	Implantable inorganic material	E0506	Hemorrhage/Blood Loss/Bleeding	A071102	Failure to Interrogate	01/06/2020	01/06/2020
							E2402	Appropriate Clinical Signs, Symptoms and Conditions Term/Code not Available	A071102	Failure to Interrogate	01/06/2020	01/06/2020
							F1905	Device Revision or Replacement	A071102	Failure to Interrogate	01/06/2020	01/06/2020
							F28	Appropriate Health Impact Term/Code not Available	A071102	Failure to Interrogate	01/06/2020	01/06/2020
					CT314	Electrically powered	E0506	Hemorrhage/Blood Loss/Bleeding	A071102	Failure to Interrogate	01/06/2020	01/06/2020
							E2402	Appropriate Clinical Signs, Symptoms and Conditions Term/Code not Available	A071102	Failure to Interrogate	01/06/2020	01/06/2020
							F1905	Device Revision or Replacement	A071102	Failure to Interrogate	01/06/2020	01/06/2020
							F28	Appropriate Health Impact Term/Code not Available	A071102	Failure to Interrogate	01/06/2020	01/06/2020
					CT321	Long-term surgical invasive	E0506	Hemorrhage/Blood Loss/Bleeding	A071102	Failure to Interrogate	01/06/2020	01/06/2020
							E2402	Appropriate Clinical Signs, Symptoms and Conditions Term/Code not Available	A071102	Failure to Interrogate	01/06/2020	01/06/2020
							F1905	Device Revision or Replacement	A071102	Failure to Interrogate	01/06/2020	01/06/2020
							F28	Appropriate Health Impact Term/Code not Available	A071102	Failure to Interrogate	01/06/2020	01/06/2020
					CT449	Generators and associated devices	E0506	Hemorrhage/Blood Loss/Bleeding	A071102	Failure to Interrogate	01/06/2020	01/06/2020
							E2402	Appropriate Clinical Signs, Symptoms and Conditions Term/Code not Available	A071102	Failure to Interrogate	01/06/2020	01/06/2020
							F1905	Device Revision or Replacement	A071102	Failure to Interrogate	01/06/2020	01/06/2020
							F28	Appropriate Health Impact Term/Code not Available	A071102	Failure to Interrogate	01/06/2020	01/06/2020
					CT981	Single use	E0506	Hemorrhage/Blood Loss/Bleeding	A071102	Failure to Interrogate	01/06/2020	01/06/2020
							E2402	Appropriate Clinical Signs, Symptoms and Conditions Term/Code not Available	A071102	Failure to Interrogate	01/06/2020	01/06/2020
							F1905	Device Revision or Replacement	A071102	Failure to Interrogate	01/06/2020	01/06/2020
							F28	Appropriate Health Impact Term/Code not Available	A071102	Failure to Interrogate	01/06/2020	01/06/2020
					CT983	Surgical invasive	E0506	Hemorrhage/Blood Loss/Bleeding	A071102	Failure to Interrogate	01/06/2020	01/06/2020
							E2402	Appropriate Clinical Signs, Symptoms and Conditions Term/Code not Available	A071102	Failure to Interrogate	01/06/2020	01/06/2020
							F1905	Device Revision or Replacement	A071102	Failure to Interrogate	01/06/2020	01/06/2020
							F28	Appropriate Health Impact Term/Code not Available	A071102	Failure to Interrogate	01/06/2020	01/06/2020
9343	2020	All other reportable incidents	36241	Pacemaker/defibrillator lead	CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A051201	Device Dislodged or Dislocated		
									A1301	Failure to Read Input Signal		
							F08	Hospitalization or Prolonged Hospitalization	A051201	Device Dislodged or Dislocated		
									A1301	Failure to Read Input Signal		
9344	2020	All other reportable incidents	47265	Dual-chamber implantable pacemaker, rate-responsive	CT141	Electrodes/Leads and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070910	Undersensing		
									A0712	Pacing Problem		
							F26	No Health Consequences or Impact	A070910	Undersensing		
									A0712	Pacing Problem		
					CT145	Cardiovascular devices	E2403	No Clinical Signs, Symptoms or Conditions	A070910	Undersensing		
									A0712	Pacing Problem		
							F26	No Health Consequences or Impact	A070910	Undersensing		
									A0712	Pacing Problem		
					CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070910	Undersensing		
									A0712	Pacing Problem		

							F26	No Health Consequences or Impact	A070910	Undersensing			
									A0712	Pacing Problem			
					CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A070910	Undersensing			
									A0712	Pacing Problem			
							F26	No Health Consequences or Impact	A070910	Undersensing			
									A0712	Pacing Problem			
					CT233	Surgical	E2403	No Clinical Signs, Symptoms or Conditions	A070910	Undersensing			
									A0712	Pacing Problem			
							F26	No Health Consequences or Impact	A070910	Undersensing			
									A0712	Pacing Problem			
					CT2406	Implantable	E2403	No Clinical Signs, Symptoms or Conditions	A070910	Undersensing			
									A0712	Pacing Problem			
							F26	No Health Consequences or Impact	A070910	Undersensing			
									A0712	Pacing Problem			
					CT2407	Active implantable	E2403	No Clinical Signs, Symptoms or Conditions	A070910	Undersensing			
									A0712	Pacing Problem			
							F26	No Health Consequences or Impact	A070910	Undersensing			
									A0712	Pacing Problem			
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A070910	Undersensing			
									A0712	Pacing Problem			
							F26	No Health Consequences or Impact	A070910	Undersensing			
									A0712	Pacing Problem			
					CT2727	Active implant devices/ancillaries	E2403	No Clinical Signs, Symptoms or Conditions	A070910	Undersensing			
									A0712	Pacing Problem			
							F26	No Health Consequences or Impact	A070910	Undersensing			
									A0712	Pacing Problem			
					CT2750	Implantable cardiac electrical stimulators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070910	Undersensing			
									A0712	Pacing Problem			
							F26	No Health Consequences or Impact	A070910	Undersensing			
									A0712	Pacing Problem			
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A070910	Undersensing			
									A0712	Pacing Problem			
							F26	No Health Consequences or Impact	A070910	Undersensing			
									A0712	Pacing Problem			
					CT2985	Active	E2403	No Clinical Signs, Symptoms or Conditions	A070910	Undersensing			
									A0712	Pacing Problem			
							F26	No Health Consequences or Impact	A070910	Undersensing			
									A0712	Pacing Problem			
					CT3001	Implantable metal	E2403	No Clinical Signs, Symptoms or Conditions	A070910	Undersensing			
									A0712	Pacing Problem			
							F26	No Health Consequences or Impact	A070910	Undersensing			
									A0712	Pacing Problem			
					CT3002	Implantable synthetic polymer	E2403	No Clinical Signs, Symptoms or Conditions	A070910	Undersensing			
									A0712	Pacing Problem			
							F26	No Health Consequences or Impact	A070910	Undersensing			
									A0712	Pacing Problem			
					CT3006	Implantable inorganic material	E2403	No Clinical Signs, Symptoms or Conditions	A070910	Undersensing			
									A0712	Pacing Problem			
							F26	No Health Consequences or Impact	A070910	Undersensing			
									A0712	Pacing Problem			
					CT314	Electrically powered	E2403	No Clinical Signs, Symptoms or Conditions	A070910	Undersensing			
									A0712	Pacing Problem			
							F26	No Health Consequences or Impact	A070910	Undersensing			
									A0712	Pacing Problem			
					CT321	Long-term surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A070910	Undersensing			
									A0712	Pacing Problem			
							F26	No Health Consequences or Impact	A070910	Undersensing			
									A0712	Pacing Problem			
					CT449	Generators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070910	Undersensing			
									A0712	Pacing Problem			
							F26	No Health Consequences or Impact	A070910	Undersensing			
									A0712	Pacing Problem			
					CT981	Single use	E2403	No Clinical Signs, Symptoms or Conditions	A070910	Undersensing			
									A0712	Pacing Problem			
							F26	No Health Consequences or Impact	A070910	Undersensing			
									A0712	Pacing Problem			
					CT983	Surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A070910	Undersensing			
									A0712	Pacing Problem			
							F26	No Health Consequences or Impact	A070910	Undersensing			
									A0712	Pacing Problem			
9345	2020	All other reportable incidents	35852	Single-chamber implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A070909	Oversensing			
							F23	Unexpected Medical Intervention	A070909	Oversensing			
9346	2019	All other reportable incidents	35852	Single-chamber implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	13/08/2019		
9347	2019	All other reportable incidents	35223	Endocardial/interventricular septal pacing lead	CT1684	Cardiac pacemakers and associated devices	F1905	Device Revision or Replacement	A072202	Low Impedance	03/07/2019		
9348	2019	All other reportable incidents	47265	Dual-chamber implantable pacemaker, rate-responsive	CT141	Electrodes/Leads and associated devices	F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	24/06/2019		
					CT145	Cardiovascular devices	F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	24/06/2019		
					CT1684	Cardiac pacemakers and associated devices	F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	24/06/2019		
					CT1685	Cardiac pacemakers	F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	24/06/2019		
					CT233	Surgical	F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	24/06/2019		
					CT2406	Implantable	F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	24/06/2019		
					CT2407	Active implantable	F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	24/06/2019		
					CT257	Implantable cardiac pacemakers	F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	24/06/2019		
					CT2727	Active implant devices/ancillaries	F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	24/06/2019		
					CT2750	Implantable cardiac electrical stimulators and associated devices	F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	24/06/2019		
					CT2866	Implantable cardiac electrical stimulators	F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	24/06/2019		
					CT2985	Active	F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	24/06/2019		

					CT3001	Implantable metal	F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	24/06/2019		
					CT3002	Implantable synthetic polymer	F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	24/06/2019		
					CT3006	Implantable inorganic material	F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	24/06/2019		
					CT314	Electrically powered	F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	24/06/2019		
					CT321	Long-term surgical invasive	F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	24/06/2019		
					CT449	Generators and associated devices	F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	24/06/2019		
					CT981	Single use	F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	24/06/2019		
					CT983	Surgical invasive	F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	24/06/2019		
9349	2019	All other reportable incidents	47265	Dual-chamber implantable pacemaker, rate-responsive	CT141	Electrodes/Leads and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070909	Oversensing			
					CT145	Cardiovascular devices	E2403	No Clinical Signs, Symptoms or Conditions	A070909	Oversensing			
					CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070909	Oversensing			
					CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A070909	Oversensing			
					CT233	Surgical	E2403	No Clinical Signs, Symptoms or Conditions	A070909	Oversensing			
					CT2406	Implantable	E2403	No Clinical Signs, Symptoms or Conditions	A070909	Oversensing			
					CT2407	Active implantable	E2403	No Clinical Signs, Symptoms or Conditions	A070909	Oversensing			
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A070909	Oversensing			
					CT2727	Active implant devices/ancillaries	E2403	No Clinical Signs, Symptoms or Conditions	A070909	Oversensing			
					CT2750	Implantable cardiac electrical stimulators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070909	Oversensing			
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A070909	Oversensing			
					CT2985	Active	E2403	No Clinical Signs, Symptoms or Conditions	A070909	Oversensing			
					CT3001	Implantable metal	E2403	No Clinical Signs, Symptoms or Conditions	A070909	Oversensing			
					CT3002	Implantable synthetic polymer	E2403	No Clinical Signs, Symptoms or Conditions	A070909	Oversensing			
					CT3006	Implantable inorganic material	E2403	No Clinical Signs, Symptoms or Conditions	A070909	Oversensing			
					CT314	Electrically powered	E2403	No Clinical Signs, Symptoms or Conditions	A070909	Oversensing			
					CT321	Long-term surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A070909	Oversensing			
					CT449	Generators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070909	Oversensing			
					CT981	Single use	E2403	No Clinical Signs, Symptoms or Conditions	A070909	Oversensing			
					CT983	Surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A070909	Oversensing			
9350	2020	All other reportable incidents	47263	Cardiac resynchronization therapy implantable pacemaker	CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070909	Oversensing			
									A072201	High Impedance			
							F23	Unexpected Medical Intervention	A070909	Oversensing			
									A072201	High Impedance			
					CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A070909	Oversensing			
									A072201	High Impedance			
							F23	Unexpected Medical Intervention	A070909	Oversensing			
									A072201	High Impedance			
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A070909	Oversensing			
									A072201	High Impedance			
							F23	Unexpected Medical Intervention	A070909	Oversensing			
									A072201	High Impedance			
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A070909	Oversensing			
									A072201	High Impedance			
							F23	Unexpected Medical Intervention	A070909	Oversensing			
									A072201	High Impedance			
9351	2020	All other reportable incidents	35852	Single-chamber implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E2104	Electric Shock	A071302	Inappropriate/Inadequate Shock/Stimulation			
							E2403	No Clinical Signs, Symptoms or Conditions	A071302	Inappropriate/Inadequate Shock/Stimulation			
							F26	No Health Consequences or Impact	A071302	Inappropriate/Inadequate Shock/Stimulation			
9352	2019	All other reportable incidents	47270	Cardiac resynchronization therapy implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	F1905	Device Revision or Replacement	A071204	No Pacing	19/07/2019		
9353	2019	All other reportable incidents	35852	Single-chamber implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A070909	Oversensing			
9354	2019	All other reportable incidents	36241	Pacemaker/defibrillator lead	CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A010402	Migration			
9355	2019	All other reportable incidents	17882	Physiologic-monitoring defibrillation system	CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070802	Intermittent Loss of Power			
							F27	Problem Identified During Non-Clinical Procedure	A070802	Intermittent Loss of Power			
9356	2019	All other reportable incidents	35223	Endocardial/interventricular septal pacing lead	CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A072201	High Impedance			
9357	2020	All other reportable incidents	35223	Endocardial/interventricular septal pacing lead	CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070101	Failure to Capture			
									A0709	Device Sensing Problem			
							F1905	Device Revision or Replacement	A070101	Failure to Capture			
									A0709	Device Sensing Problem			
9358	2019	All other reportable incidents	47263	Cardiac resynchronization therapy implantable pacemaker	CT1684	Cardiac pacemakers and associated devices	F1905	Device Revision or Replacement	A	Annex A	09/07/2019		
					CT1685	Cardiac pacemakers	F1905	Device Revision or Replacement	A	Annex A	09/07/2019		
					CT257	Implantable cardiac pacemakers	F1905	Device Revision or Replacement	A	Annex A	09/07/2019		
					CT2866	Implantable cardiac electrical stimulators	F1905	Device Revision or Replacement	A	Annex A	09/07/2019		
9359	2019	All other reportable incidents	35223	Endocardial/interventricular septal pacing lead	CT1684	Cardiac pacemakers and associated devices	F1905	Device Revision or Replacement	A070909	Oversensing			
9360	2019	All other reportable incidents	47265	Dual-chamber implantable pacemaker, rate-responsive	CT141	Electrodes/Leads and associated devices	F1905	Device Revision or Replacement	A071102	Failure to Interrogate	05/07/2019		
					CT145	Cardiovascular devices	F1905	Device Revision or Replacement	A071102	Failure to Interrogate	05/07/2019		
					CT1684	Cardiac pacemakers and associated devices	F1905	Device Revision or Replacement	A071102	Failure to Interrogate	05/07/2019		
					CT1685	Cardiac pacemakers	F1905	Device Revision or Replacement	A071102	Failure to Interrogate	05/07/2019		
					CT233	Surgical	F1905	Device Revision or Replacement	A071102	Failure to Interrogate	05/07/2019		
					CT2406	Implantable	F1905	Device Revision or Replacement	A071102	Failure to Interrogate	05/07/2019		
					CT2407	Active implantable	F1905	Device Revision or Replacement	A071102	Failure to Interrogate	05/07/2019		
					CT257	Implantable cardiac pacemakers	F1905	Device Revision or Replacement	A071102	Failure to Interrogate	05/07/2019		
					CT2727	Active implant devices/ancillaries	F1905	Device Revision or Replacement	A071102	Failure to Interrogate	05/07/2019		
					CT2750	Implantable cardiac electrical stimulators and associated devices	F1905	Device Revision or Replacement	A071102	Failure to Interrogate	05/07/2019		
					CT2866	Implantable cardiac electrical stimulators	F1905	Device Revision or Replacement	A071102	Failure to Interrogate	05/07/2019		
					CT2985	Active	F1905	Device Revision or Replacement	A071102	Failure to Interrogate	05/07/2019		
					CT3001	Implantable metal	F1905	Device Revision or Replacement	A071102	Failure to Interrogate	05/07/2019		
					CT3002	Implantable synthetic polymer	F1905	Device Revision or Replacement	A071102	Failure to Interrogate	05/07/2019		
					CT3006	Implantable inorganic material	F1905	Device Revision or Replacement	A071102	Failure to Interrogate	05/07/2019		
					CT314	Electrically powered	F1905	Device Revision or Replacement	A071102	Failure to Interrogate	05/07/2019		
					CT321	Long-term surgical invasive	F1905	Device Revision or Replacement	A071102	Failure to Interrogate	05/07/2019		
					CT449	Generators and associated devices	F1905	Device Revision or Replacement	A071102	Failure to Interrogate	05/07/2019		
					CT981	Single use	F1905	Device Revision or Replacement	A071102	Failure to Interrogate	05/07/2019		
					CT983	Surgical invasive	F1905	Device Revision or Replacement	A071102	Failure to Interrogate	05/07/2019		
9361	2019	All other reportable incidents	37265	Dual-chamber implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A0721	Circuit Failure			
9362	2019	All other reportable incidents	17882	Physiologic-monitoring defibrillation system	CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A0721	Circuit Failure			

9363	2019	All other reportable incidents	37265	Dual-chamber implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	17/07/2019	
9364	2020	All other reportable incidents	47265	Dual-chamber implantable pacemaker, rate-responsive	CT141	Electrodes/Leads and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	28/05/2020	28/05/2020
							F19	Surgical Intervention	A070504	Premature Discharge of Battery	28/05/2020	28/05/2020
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	28/05/2020	28/05/2020
					CT145	Cardiovascular devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	28/05/2020	28/05/2020
							F19	Surgical Intervention	A070504	Premature Discharge of Battery	28/05/2020	28/05/2020
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	28/05/2020	28/05/2020
					CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	28/05/2020	28/05/2020
							F19	Surgical Intervention	A070504	Premature Discharge of Battery	28/05/2020	28/05/2020
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	28/05/2020	28/05/2020
					CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	28/05/2020	28/05/2020
							F19	Surgical Intervention	A070504	Premature Discharge of Battery	28/05/2020	28/05/2020
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	28/05/2020	28/05/2020
					CT233	Surgical	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	28/05/2020	28/05/2020
							F19	Surgical Intervention	A070504	Premature Discharge of Battery	28/05/2020	28/05/2020
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	28/05/2020	28/05/2020
					CT2406	Implantable	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	28/05/2020	28/05/2020
							F19	Surgical Intervention	A070504	Premature Discharge of Battery	28/05/2020	28/05/2020
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	28/05/2020	28/05/2020
					CT2407	Active implantable	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	28/05/2020	28/05/2020
							F19	Surgical Intervention	A070504	Premature Discharge of Battery	28/05/2020	28/05/2020
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	28/05/2020	28/05/2020
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	28/05/2020	28/05/2020
							F19	Surgical Intervention	A070504	Premature Discharge of Battery	28/05/2020	28/05/2020
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	28/05/2020	28/05/2020
					CT2727	Active implant devices/ancillaries	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	28/05/2020	28/05/2020
							F19	Surgical Intervention	A070504	Premature Discharge of Battery	28/05/2020	28/05/2020
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	28/05/2020	28/05/2020
					CT2750	Implantable cardiac electrical stimulators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	28/05/2020	28/05/2020
							F19	Surgical Intervention	A070504	Premature Discharge of Battery	28/05/2020	28/05/2020
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	28/05/2020	28/05/2020
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	28/05/2020	28/05/2020
							F19	Surgical Intervention	A070504	Premature Discharge of Battery	28/05/2020	28/05/2020
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	28/05/2020	28/05/2020
					CT2985	Active	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	28/05/2020	28/05/2020
							F19	Surgical Intervention	A070504	Premature Discharge of Battery	28/05/2020	28/05/2020
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	28/05/2020	28/05/2020
					CT3001	Implantable metal	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	28/05/2020	28/05/2020
							F19	Surgical Intervention	A070504	Premature Discharge of Battery	28/05/2020	28/05/2020
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	28/05/2020	28/05/2020
					CT3002	Implantable synthetic polymer	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	28/05/2020	28/05/2020
							F19	Surgical Intervention	A070504	Premature Discharge of Battery	28/05/2020	28/05/2020
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	28/05/2020	28/05/2020
					CT3006	Implantable inorganic material	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	28/05/2020	28/05/2020
							F19	Surgical Intervention	A070504	Premature Discharge of Battery	28/05/2020	28/05/2020
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	28/05/2020	28/05/2020
					CT314	Electrically powered	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	28/05/2020	28/05/2020
							F19	Surgical Intervention	A070504	Premature Discharge of Battery	28/05/2020	28/05/2020
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	28/05/2020	28/05/2020
					CT321	Long-term surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	28/05/2020	28/05/2020
							F19	Surgical Intervention	A070504	Premature Discharge of Battery	28/05/2020	28/05/2020
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	28/05/2020	28/05/2020
					CT449	Generators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	28/05/2020	28/05/2020
							F19	Surgical Intervention	A070504	Premature Discharge of Battery	28/05/2020	28/05/2020
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	28/05/2020	28/05/2020
					CT981	Single use	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	28/05/2020	28/05/2020
							F19	Surgical Intervention	A070504	Premature Discharge of Battery	28/05/2020	28/05/2020
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	28/05/2020	28/05/2020
					CT983	Surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	28/05/2020	28/05/2020
							F19	Surgical Intervention	A070504	Premature Discharge of Battery	28/05/2020	28/05/2020
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	28/05/2020	28/05/2020
9365	2019	All other reportable incidents	47265	Dual-chamber implantable pacemaker, rate-responsive	CT141	Electrodes/Leads and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A0722	Impedance Problem		
					CT145	Cardiovascular devices	E2403	No Clinical Signs, Symptoms or Conditions	A0722	Impedance Problem		
					CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A0722	Impedance Problem		
					CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A0722	Impedance Problem		
					CT233	Surgical	E2403	No Clinical Signs, Symptoms or Conditions	A0722	Impedance Problem		
					CT2406	Implantable	E2403	No Clinical Signs, Symptoms or Conditions	A0722	Impedance Problem		
					CT2407	Active implantable	E2403	No Clinical Signs, Symptoms or Conditions	A0722	Impedance Problem		
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A0722	Impedance Problem		
					CT2727	Active implant devices/ancillaries	E2403	No Clinical Signs, Symptoms or Conditions	A0722	Impedance Problem		
					CT2750	Implantable cardiac electrical stimulators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A0722	Impedance Problem		
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A0722	Impedance Problem		
					CT2985	Active	E2403	No Clinical Signs, Symptoms or Conditions	A0722	Impedance Problem		
					CT3001	Implantable metal	E2403	No Clinical Signs, Symptoms or Conditions	A0722	Impedance Problem		
					CT3002	Implantable synthetic polymer	E2403	No Clinical Signs, Symptoms or Conditions	A0722	Impedance Problem		
					CT3006	Implantable inorganic material	E2403	No Clinical Signs, Symptoms or Conditions	A0722	Impedance Problem		
					CT314	Electrically powered	E2403	No Clinical Signs, Symptoms or Conditions	A0722	Impedance Problem		
					CT321	Long-term surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A0722	Impedance Problem		
					CT449	Generators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A0722	Impedance Problem		
					CT981	Single use	E2403	No Clinical Signs, Symptoms or Conditions	A0722	Impedance Problem		
					CT983	Surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A0722	Impedance Problem		
9366	2019	All other reportable incidents	36241	Pacemaker/defibrillator lead	CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A0722	Impedance Problem		
9367	2020	All other reportable incidents	35852	Single-chamber implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E2104	Electric Shock	A070909	Oversensing		
									A071302	Inappropriate/Inadequate Shock/Stimulation		
							F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A070909	Oversensing		
									A071302	Inappropriate/Inadequate Shock/Stimulation		
							F15	Recognized Device or Procedural Complication	A070909	Oversensing		

									A071302	Inappropriate/Inadequate Shock/Stimulation		
9368	2019	All other reportable incidents	35852	Single-chamber implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery		
9369	2019	All other reportable incidents	17882	Physiologic-monitoring defibrillation system	CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070802	Intermittent Loss of Power		
9370	2020	All other reportable incidents	36241	Pacemaker/defibrillator lead	CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070101	Failure to Capture		
							F26	No Health Consequences or Impact	A070101	Failure to Capture		
9371	2020	All other reportable incidents	47263	Cardiac resynchronization therapy implantable pacemaker	CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	08/06/2020	08/06/2020
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	08/06/2020	08/06/2020
					CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	08/06/2020	08/06/2020
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	08/06/2020	08/06/2020
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	08/06/2020	08/06/2020
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	08/06/2020	08/06/2020
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	08/06/2020	08/06/2020
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	08/06/2020	08/06/2020
9372	2020	All other reportable incidents	35223	Endocardial/interventricular septal pacing lead	CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A072201	High Impedance		
							F26	No Health Consequences or Impact	A072201	High Impedance		
9373	2020	All other reportable incidents	17882	Physiologic-monitoring defibrillation system	CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A071301	Failure to Deliver Shock/Stimulation		
							F27	Problem Identified During Non-Clinical Procedure	A071301	Failure to Deliver Shock/Stimulation		
9374	2020	All other reportable incidents	47267	Single-chamber implantable pacemaker, rate-responsive	CT141	Electrodes/Leads and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT145	Cardiovascular devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT233	Surgical	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT2406	Implantable	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT2407	Active implantable	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT2727	Active implant devices/ancillaries	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT2750	Implantable cardiac electrical stimulators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT2985	Active	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT3001	Implantable metal	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT3002	Implantable synthetic polymer	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT3006	Implantable inorganic material	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT314	Electrically powered	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT321	Long-term surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT449	Generators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT981	Single use	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT983	Surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
9375	2019	All other reportable incidents	47267	Single-chamber implantable pacemaker, rate-responsive	CT141	Electrodes/Leads and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	13/07/2019	
					CT145	Cardiovascular devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	13/07/2019	
					CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	13/07/2019	
					CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	13/07/2019	
					CT233	Surgical	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	13/07/2019	
					CT2406	Implantable	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	13/07/2019	
					CT2407	Active implantable	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	13/07/2019	
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	13/07/2019	
					CT2727	Active implant devices/ancillaries	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	13/07/2019	
					CT2750	Implantable cardiac electrical stimulators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	13/07/2019	

					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	13/07/2019		
					CT2985	Active	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	13/07/2019		
					CT3001	Implantable metal	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	13/07/2019		
					CT3002	Implantable synthetic polymer	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	13/07/2019		
					CT3006	Implantable inorganic material	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	13/07/2019		
					CT314	Electrically powered	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	13/07/2019		
					CT321	Long-term surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	13/07/2019		
					CT449	Generators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	13/07/2019		
					CT981	Single use	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	13/07/2019		
					CT983	Surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	13/07/2019		
9376	2019	All other reportable incidents	47263	Cardiac resynchronization therapy implantable pacemaker	CT1684	Cardiac pacemakers and associated devices	E1906	Unspecified Infection	A010102	Device Appears to Trigger Rejection			
							F1905	Device Revision or Replacement	A010102	Device Appears to Trigger Rejection			
					CT1685	Cardiac pacemakers	E1906	Unspecified Infection	A010102	Device Appears to Trigger Rejection			
							F1905	Device Revision or Replacement	A010102	Device Appears to Trigger Rejection			
					CT257	Implantable cardiac pacemakers	E1906	Unspecified Infection	A010102	Device Appears to Trigger Rejection			
							F1905	Device Revision or Replacement	A010102	Device Appears to Trigger Rejection			
					CT2866	Implantable cardiac electrical stimulators	E1906	Unspecified Infection	A010102	Device Appears to Trigger Rejection			
							F1905	Device Revision or Replacement	A010102	Device Appears to Trigger Rejection			
9377	2019	All other reportable incidents	47270	Cardiac resynchronization therapy implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A070909	Oversensing	03/02/2021	03/02/2021	
									A072201	High Impedance	03/02/2021	03/02/2021	
							F26	No Health Consequences or Impact	A070909	Oversensing	03/02/2021	03/02/2021	
									A072201	High Impedance	03/02/2021	03/02/2021	
9378	2019	All other reportable incidents	36241	Pacemaker/defibrillator lead	CT1684	Cardiac pacemakers and associated devices	F1905	Device Revision or Replacement	A072201	High Impedance	01/07/2019		
9379	2019	All other reportable incidents	47267	Single-chamber implantable pacemaker, rate-responsive	CT141	Electrodes/Leads and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery			
					CT145	Cardiovascular devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery			
					CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery			
					CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery			
					CT233	Surgical	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery			
					CT2406	Implantable	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery			
					CT2407	Active implantable	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery			
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery			
					CT2727	Active implant devices/ancillaries	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery			
					CT2750	Implantable cardiac electrical stimulators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery			
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery			
					CT2985	Active	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery			
					CT3001	Implantable metal	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery			
					CT3002	Implantable synthetic polymer	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery			
					CT3006	Implantable inorganic material	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery			
					CT314	Electrically powered	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery			
					CT321	Long-term surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery			
					CT449	Generators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery			
					CT981	Single use	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery			
					CT983	Surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery			
9380	2019	All other reportable incidents	36241	Pacemaker/defibrillator lead	CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A150206	Difficult to Insert	01/07/2019	01/07/2019	
9381	2019	All other reportable incidents	36241	Pacemaker/defibrillator lead	CT1684	Cardiac pacemakers and associated devices	F1905	Device Revision or Replacement	A051201	Device Dislodged or Dislocated	20/02/2019	20/02/2019	
9382	2019	All other reportable incidents	47265	Dual-chamber implantable pacemaker, rate-responsive	CT141	Electrodes/Leads and associated devices	F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	29/07/2019		
					CT145	Cardiovascular devices	F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	29/07/2019		
					CT1684	Cardiac pacemakers and associated devices	F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	29/07/2019		
					CT1685	Cardiac pacemakers	F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	29/07/2019		
					CT233	Surgical	F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	29/07/2019		
					CT2406	Implantable	F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	29/07/2019		
					CT2407	Active implantable	F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	29/07/2019		
					CT257	Implantable cardiac pacemakers	F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	29/07/2019		
					CT2727	Active implant devices/ancillaries	F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	29/07/2019		
					CT2750	Implantable cardiac electrical stimulators and associated devices	F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	29/07/2019		
					CT2866	Implantable cardiac electrical stimulators	F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	29/07/2019		
					CT2985	Active	F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	29/07/2019		
					CT3001	Implantable metal	F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	29/07/2019		
					CT3002	Implantable synthetic polymer	F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	29/07/2019		
					CT3006	Implantable inorganic material	F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	29/07/2019		
					CT314	Electrically powered	F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	29/07/2019		
					CT321	Long-term surgical invasive	F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	29/07/2019		
					CT449	Generators and associated devices	F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	29/07/2019		
					CT981	Single use	F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	29/07/2019		
					CT983	Surgical invasive	F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	29/07/2019		
9383	2019	All other reportable incidents	35852	Single-chamber implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A070909	Oversensing			
9384	2019	All other reportable incidents	47265	Dual-chamber implantable pacemaker, rate-responsive	CT141	Electrodes/Leads and associated devices	F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A070101	Failure to Capture			
					CT145	Cardiovascular devices	F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A070101	Failure to Capture			
					CT1684	Cardiac pacemakers and associated devices	F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A070101	Failure to Capture			
					CT1685	Cardiac pacemakers	F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A070101	Failure to Capture			
					CT233	Surgical	F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A070101	Failure to Capture			
					CT2406	Implantable	F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A070101	Failure to Capture			
					CT2407	Active implantable	F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A070101	Failure to Capture			
					CT257	Implantable cardiac pacemakers	F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A070101	Failure to Capture			
					CT2727	Active implant devices/ancillaries	F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A070101	Failure to Capture			
					CT2750	Implantable cardiac electrical stimulators and associated devices	F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A070101	Failure to Capture			
					CT2866	Implantable cardiac electrical stimulators	F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A070101	Failure to Capture			
					CT2985	Active	F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A070101	Failure to Capture			
					CT3001	Implantable metal	F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A070101	Failure to Capture			
					CT3002	Implantable synthetic polymer	F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A070101	Failure to Capture			
					CT3006	Implantable inorganic material	F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A070101	Failure to Capture			
					CT314	Electrically powered	F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A070101	Failure to Capture			
					CT321	Long-term surgical invasive	F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A070101	Failure to Capture			
					CT449	Generators and associated devices	F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A070101	Failure to Capture			
					CT981	Single use	F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A070101	Failure to Capture			

					CT983	Surgical invasive	F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A070101	Failure to Capture		
9385	2019	All other reportable incidents	35852	Single-chamber implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A070909	Oversensing		
9386	2019	Death	47265	Dual-chamber implantable pacemaker, rate-responsive	CT141	Electrodes/Leads and associated devices	F21	Unexpected Deterioration	A26	Insufficient Device Problem Information		
					CT145	Cardiovascular devices	F21	Unexpected Deterioration	A26	Insufficient Device Problem Information		
					CT1684	Cardiac pacemakers and associated devices	F21	Unexpected Deterioration	A26	Insufficient Device Problem Information		
					CT1685	Cardiac pacemakers	F21	Unexpected Deterioration	A26	Insufficient Device Problem Information		
					CT233	Surgical	F21	Unexpected Deterioration	A26	Insufficient Device Problem Information		
					CT2406	Implantable	F21	Unexpected Deterioration	A26	Insufficient Device Problem Information		
					CT2407	Active implantable	F21	Unexpected Deterioration	A26	Insufficient Device Problem Information		
					CT257	Implantable cardiac pacemakers	F21	Unexpected Deterioration	A26	Insufficient Device Problem Information		
					CT2727	Active implant devices/ancillaries	F21	Unexpected Deterioration	A26	Insufficient Device Problem Information		
					CT2750	Implantable cardiac electrical stimulators and associated devices	F21	Unexpected Deterioration	A26	Insufficient Device Problem Information		
					CT2866	Implantable cardiac electrical stimulators	F21	Unexpected Deterioration	A26	Insufficient Device Problem Information		
					CT2985	Active	F21	Unexpected Deterioration	A26	Insufficient Device Problem Information		
					CT3001	Implantable metal	F21	Unexpected Deterioration	A26	Insufficient Device Problem Information		
					CT3002	Implantable synthetic polymer	F21	Unexpected Deterioration	A26	Insufficient Device Problem Information		
					CT3006	Implantable inorganic material	F21	Unexpected Deterioration	A26	Insufficient Device Problem Information		
					CT314	Electrically powered	F21	Unexpected Deterioration	A26	Insufficient Device Problem Information		
					CT321	Long-term surgical invasive	F21	Unexpected Deterioration	A26	Insufficient Device Problem Information		
					CT449	Generators and associated devices	F21	Unexpected Deterioration	A26	Insufficient Device Problem Information		
					CT981	Single use	F21	Unexpected Deterioration	A26	Insufficient Device Problem Information		
					CT983	Surgical invasive	F21	Unexpected Deterioration	A26	Insufficient Device Problem Information		
9387	2019	All other reportable incidents	37265	Dual-chamber implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	08/07/2019	
9388	2019	All other reportable incidents	35852	Single-chamber implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	09/07/2019	
9389	2019	All other reportable incidents	47265	Dual-chamber implantable pacemaker, rate-responsive	CT141	Electrodes/Leads and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
					CT145	Cardiovascular devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
					CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
					CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
					CT233	Surgical	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
					CT2406	Implantable	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
					CT2407	Active implantable	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
					CT2727	Active implant devices/ancillaries	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
					CT2750	Implantable cardiac electrical stimulators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
					CT2985	Active	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
					CT3001	Implantable metal	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
					CT3002	Implantable synthetic polymer	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
					CT3006	Implantable inorganic material	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
					CT314	Electrically powered	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
					CT321	Long-term surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
					CT449	Generators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
					CT981	Single use	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
					CT983	Surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
9390	2019	All other reportable incidents	47265	Dual-chamber implantable pacemaker, rate-responsive	CT141	Electrodes/Leads and associated devices	F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	12/07/2019	
					CT145	Cardiovascular devices	F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	12/07/2019	
					CT1684	Cardiac pacemakers and associated devices	F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	12/07/2019	
					CT1685	Cardiac pacemakers	F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	12/07/2019	
					CT233	Surgical	F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	12/07/2019	
					CT2406	Implantable	F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	12/07/2019	
					CT2407	Active implantable	F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	12/07/2019	
					CT257	Implantable cardiac pacemakers	F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	12/07/2019	
					CT2727	Active implant devices/ancillaries	F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	12/07/2019	
					CT2750	Implantable cardiac electrical stimulators and associated devices	F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	12/07/2019	
					CT2866	Implantable cardiac electrical stimulators	F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	12/07/2019	
					CT2985	Active	F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	12/07/2019	
					CT3001	Implantable metal	F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	12/07/2019	
					CT3002	Implantable synthetic polymer	F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	12/07/2019	
					CT3006	Implantable inorganic material	F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	12/07/2019	
					CT314	Electrically powered	F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	12/07/2019	
					CT321	Long-term surgical invasive	F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	12/07/2019	
					CT449	Generators and associated devices	F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	12/07/2019	
					CT981	Single use	F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	12/07/2019	
					CT983	Surgical invasive	F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	12/07/2019	
9391	2019	All other reportable incidents	47265	Dual-chamber implantable pacemaker, rate-responsive	CT141	Electrodes/Leads and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
					CT145	Cardiovascular devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
					CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
					CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
					CT233	Surgical	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
					CT2406	Implantable	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
					CT2407	Active implantable	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
					CT2727	Active implant devices/ancillaries	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
					CT2750	Implantable cardiac electrical stimulators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
					CT2985	Active	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
					CT3001	Implantable metal	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
					CT3002	Implantable synthetic polymer	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
					CT3006	Implantable inorganic material	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
					CT314	Electrically powered	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
					CT321	Long-term surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
					CT449	Generators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
					CT981	Single use	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
					CT983	Surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
9392	2019	All other reportable incidents	47267	Single-chamber implantable pacemaker, rate-	CT141	Electrodes/Leads and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		

				responsive	CT145	Cardiovascular devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
					CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
					CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
					CT233	Surgical	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
					CT2406	Implantable	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
					CT2407	Active implantable	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
					CT2727	Active implant devices/ancillaries	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
					CT2750	Implantable cardiac electrical stimulators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
					CT2985	Active	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
					CT3001	Implantable metal	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
					CT3002	Implantable synthetic polymer	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
					CT3006	Implantable inorganic material	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
					CT314	Electrically powered	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
					CT321	Long-term surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
					CT449	Generators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
					CT981	Single use	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
					CT983	Surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
9393	2020	All other reportable incidents	35852	Single-chamber implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E2104	Electric Shock	A070909	Oversensing		
									A071302	Inappropriate/Inadequate Shock/Stimulation		
							F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A070909	Oversensing		
									A071302	Inappropriate/Inadequate Shock/Stimulation		
							F11	Non-Serious Injury/ Illness/ Impairment	A070909	Oversensing		
									A071302	Inappropriate/Inadequate Shock/Stimulation		
9394	2020	All other reportable incidents	35852	Single-chamber implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E2104	Electric Shock	A071302	Inappropriate/Inadequate Shock/Stimulation		
							F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A071302	Inappropriate/Inadequate Shock/Stimulation		
							F26	No Health Consequences or Impact	A071302	Inappropriate/Inadequate Shock/Stimulation		
9395	2020	All other reportable incidents	47270	Cardiac resynchronization therapy implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A072201	High Impedance		
							F26	No Health Consequences or Impact	A072201	High Impedance		
9396	2019	All other reportable incidents	60789	Intracardiac pacemaker	CT130	Catheters and associated devices	F1905	Device Revision or Replacement	A1305	Wireless Communication Problem		
					CT145	Cardiovascular devices	F1905	Device Revision or Replacement	A1305	Wireless Communication Problem		
					CT1684	Cardiac pacemakers and associated devices	F1905	Device Revision or Replacement	A1305	Wireless Communication Problem		
					CT1685	Cardiac pacemakers	F1905	Device Revision or Replacement	A1305	Wireless Communication Problem		
					CT233	Surgical	F1905	Device Revision or Replacement	A1305	Wireless Communication Problem		
					CT2406	Implantable	F1905	Device Revision or Replacement	A1305	Wireless Communication Problem		
					CT2407	Active implantable	F1905	Device Revision or Replacement	A1305	Wireless Communication Problem		
					CT257	Implantable cardiac pacemakers	F1905	Device Revision or Replacement	A1305	Wireless Communication Problem		
					CT2727	Active implant devices/ancillaries	F1905	Device Revision or Replacement	A1305	Wireless Communication Problem		
					CT2750	Implantable cardiac electrical stimulators and associated devices	F1905	Device Revision or Replacement	A1305	Wireless Communication Problem		
					CT2818	Catheterization/Cannulation/Fluid path devices	F1905	Device Revision or Replacement	A1305	Wireless Communication Problem		
					CT2866	Implantable cardiac electrical stimulators	F1905	Device Revision or Replacement	A1305	Wireless Communication Problem		
					CT2985	Active	F1905	Device Revision or Replacement	A1305	Wireless Communication Problem		
					CT3001	Implantable metal	F1905	Device Revision or Replacement	A1305	Wireless Communication Problem		
					CT3002	Implantable synthetic polymer	F1905	Device Revision or Replacement	A1305	Wireless Communication Problem		
					CT3006	Implantable inorganic material	F1905	Device Revision or Replacement	A1305	Wireless Communication Problem		
					CT314	Electrically powered	F1905	Device Revision or Replacement	A1305	Wireless Communication Problem		
					CT321	Long-term surgical invasive	F1905	Device Revision or Replacement	A1305	Wireless Communication Problem		
					CT981	Single use	F1905	Device Revision or Replacement	A1305	Wireless Communication Problem		
					CT983	Surgical invasive	F1905	Device Revision or Replacement	A1305	Wireless Communication Problem		
9397	2019	All other reportable incidents	47267	Single-chamber implantable pacemaker, rate-responsive	CT141	Electrodes/Leads and associated devices	F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	05/07/2019	
					CT145	Cardiovascular devices	F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	05/07/2019	
					CT1684	Cardiac pacemakers and associated devices	F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	05/07/2019	
					CT1685	Cardiac pacemakers	F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	05/07/2019	
					CT233	Surgical	F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	05/07/2019	
					CT2406	Implantable	F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	05/07/2019	
					CT2407	Active implantable	F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	05/07/2019	
					CT257	Implantable cardiac pacemakers	F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	05/07/2019	
					CT2727	Active implant devices/ancillaries	F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	05/07/2019	
					CT2750	Implantable cardiac electrical stimulators and associated devices	F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	05/07/2019	
					CT2866	Implantable cardiac electrical stimulators	F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	05/07/2019	
					CT2985	Active	F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	05/07/2019	
					CT3001	Implantable metal	F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	05/07/2019	
					CT3002	Implantable synthetic polymer	F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	05/07/2019	
					CT3006	Implantable inorganic material	F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	05/07/2019	
					CT314	Electrically powered	F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	05/07/2019	
					CT321	Long-term surgical invasive	F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	05/07/2019	
					CT449	Generators and associated devices	F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	05/07/2019	
					CT981	Single use	F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	05/07/2019	
					CT983	Surgical invasive	F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	05/07/2019	
9398	2019	All other reportable incidents	35852	Single-chamber implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A070907	Loss of Threshold		
9399	2019	All other reportable incidents	47270	Cardiac resynchronization therapy implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A071302	Inappropriate/Inadequate Shock/Stimulation		
9400	2019	All other reportable incidents	35852	Single-chamber implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A070909	Oversensing	03/07/2019	03/07/2019
							F1908	Prolonged Surgery	A070909	Oversensing	03/07/2019	03/07/2019
9401	2019	All other reportable incidents	35852	Single-chamber implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E2104	Electric Shock	A070909	Oversensing	10/01/2020	10/01/2020
									A071302	Inappropriate/Inadequate Shock/Stimulation	10/01/2020	10/01/2020
							E2403	No Clinical Signs, Symptoms or Conditions	A070909	Oversensing	10/01/2020	10/01/2020
									A071302	Inappropriate/Inadequate Shock/Stimulation	10/01/2020	10/01/2020
							F08	Hospitalization or Prolonged Hospitalization	A070909	Oversensing	10/01/2020	10/01/2020
									A071302	Inappropriate/Inadequate Shock/Stimulation	10/01/2020	10/01/2020
							F12	Serious Injury/ Illness/ Impairment	A070909	Oversensing	10/01/2020	10/01/2020
									A071302	Inappropriate/Inadequate Shock/Stimulation	10/01/2020	10/01/2020
							F1905	Device Revision or Replacement	A070909	Oversensing	10/01/2020	10/01/2020
									A071302	Inappropriate/Inadequate Shock/Stimulation	10/01/2020	10/01/2020
9402	2019	All other reportable incidents	37265	Dual-chamber implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	30/08/2019	30/08/2019

9403	2020	All other reportable incidents	35852	Single-chamber implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E2104	Electric Shock	A070909	Oversensing		
									A071302	Inappropriate/Inadequate Shock/Stimulation		
							E2403	No Clinical Signs, Symptoms or Conditions	A070909	Oversensing		
									A071302	Inappropriate/Inadequate Shock/Stimulation		
							F26	No Health Consequences or Impact	A070909	Oversensing		
									A071302	Inappropriate/Inadequate Shock/Stimulation		
9404	2020	All other reportable incidents	35852	Single-chamber implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E2104	Electric Shock	A0205	Packaging Problem	19/05/2020	19/05/2020
									A070909	Oversensing	19/05/2020	19/05/2020
									A071302	Inappropriate/Inadequate Shock/Stimulation	19/05/2020	19/05/2020
									A072201	High Impedance	19/05/2020	19/05/2020
							F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A0205	Packaging Problem	19/05/2020	19/05/2020
									A070909	Oversensing	19/05/2020	19/05/2020
									A071302	Inappropriate/Inadequate Shock/Stimulation	19/05/2020	19/05/2020
									A072201	High Impedance	19/05/2020	19/05/2020
							F12	Serious Injury/ Illness/ Impairment	A0205	Packaging Problem	19/05/2020	19/05/2020
									A070909	Oversensing	19/05/2020	19/05/2020
									A071302	Inappropriate/Inadequate Shock/Stimulation	19/05/2020	19/05/2020
									A072201	High Impedance	19/05/2020	19/05/2020
							F19	Surgical Intervention	A0205	Packaging Problem	19/05/2020	19/05/2020
									A070909	Oversensing	19/05/2020	19/05/2020
									A071302	Inappropriate/Inadequate Shock/Stimulation	19/05/2020	19/05/2020
									A072201	High Impedance	19/05/2020	19/05/2020
							F1905	Device Revision or Replacement	A0205	Packaging Problem	19/05/2020	19/05/2020
									A070909	Oversensing	19/05/2020	19/05/2020
									A071302	Inappropriate/Inadequate Shock/Stimulation	19/05/2020	19/05/2020
									A072201	High Impedance	19/05/2020	19/05/2020
9405	2020	All other reportable incidents	36241	Pacemaker/defibrillator lead	CT1684	Cardiac pacemakers and associated devices	E2401	Insufficient Information	A070101	Failure to Capture		
									A070102	High Capture Threshold		
							E2403	No Clinical Signs, Symptoms or Conditions	A070101	Failure to Capture		
									A070102	High Capture Threshold		
							F26	No Health Consequences or Impact	A070101	Failure to Capture		
									A070102	High Capture Threshold		
9406	2020	All other reportable incidents	35852	Single-chamber implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E060109	Tachycardia	A071202	Inaccurate Synchronization		
							E2403	No Clinical Signs, Symptoms or Conditions	A071202	Inaccurate Synchronization		
							F26	No Health Consequences or Impact	A071202	Inaccurate Synchronization		
9407	2019	All other reportable incidents	35852	Single-chamber implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	01/08/2019	
9408	2019	All other reportable incidents	47270	Cardiac resynchronization therapy implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	15/07/2019	
9409	2020	All other reportable incidents	17882	Physiologic-monitoring defibrillation system	CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A07	Electrical /Electronic Property Problem		
							F27	Problem Identified During Non-Clinical Procedure	A07	Electrical /Electronic Property Problem		
9410	2019	All other reportable incidents	35223	Endocardial/interventricular septal pacing lead	CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070909	Oversensing	12/07/2019	
9411	2019	All other reportable incidents	47267	Single-chamber implantable pacemaker, rate-responsive	CT141	Electrodes/Leads and associated devices	E0112	Dizziness	A070909	Oversensing		
							E0752	Unspecified Respiratory Problem	A070909	Oversensing		
					CT145	Cardiovascular devices	E0112	Dizziness	A070909	Oversensing		
							E0752	Unspecified Respiratory Problem	A070909	Oversensing		
					CT1684	Cardiac pacemakers and associated devices	E0112	Dizziness	A070909	Oversensing		
							E0752	Unspecified Respiratory Problem	A070909	Oversensing		
					CT1685	Cardiac pacemakers	E0112	Dizziness	A070909	Oversensing		
							E0752	Unspecified Respiratory Problem	A070909	Oversensing		
					CT233	Surgical	E0112	Dizziness	A070909	Oversensing		
							E0752	Unspecified Respiratory Problem	A070909	Oversensing		
					CT2406	Implantable	E0112	Dizziness	A070909	Oversensing		
							E0752	Unspecified Respiratory Problem	A070909	Oversensing		
					CT2407	Active implantable	E0112	Dizziness	A070909	Oversensing		
							E0752	Unspecified Respiratory Problem	A070909	Oversensing		
					CT257	Implantable cardiac pacemakers	E0112	Dizziness	A070909	Oversensing		
							E0752	Unspecified Respiratory Problem	A070909	Oversensing		
					CT2727	Active implant devices/ancillaries	E0112	Dizziness	A070909	Oversensing		
							E0752	Unspecified Respiratory Problem	A070909	Oversensing		
					CT2750	Implantable cardiac electrical stimulators and associated devices	E0112	Dizziness	A070909	Oversensing		
							E0752	Unspecified Respiratory Problem	A070909	Oversensing		
					CT2866	Implantable cardiac electrical stimulators	E0112	Dizziness	A070909	Oversensing		
							E0752	Unspecified Respiratory Problem	A070909	Oversensing		
					CT2985	Active	E0112	Dizziness	A070909	Oversensing		
							E0752	Unspecified Respiratory Problem	A070909	Oversensing		
					CT3001	Implantable metal	E0112	Dizziness	A070909	Oversensing		
							E0752	Unspecified Respiratory Problem	A070909	Oversensing		
					CT3002	Implantable synthetic polymer	E0112	Dizziness	A070909	Oversensing		
							E0752	Unspecified Respiratory Problem	A070909	Oversensing		
					CT3006	Implantable inorganic material	E0112	Dizziness	A070909	Oversensing		
							E0752	Unspecified Respiratory Problem	A070909	Oversensing		
					CT314	Electrically powered	E0112	Dizziness	A070909	Oversensing		
							E0752	Unspecified Respiratory Problem	A070909	Oversensing		
					CT321	Long-term surgical invasive	E0112	Dizziness	A070909	Oversensing		
							E0752	Unspecified Respiratory Problem	A070909	Oversensing		
					CT449	Generators and associated devices	E0112	Dizziness	A070909	Oversensing		
							E0752	Unspecified Respiratory Problem	A070909	Oversensing		
					CT981	Single use	E0112	Dizziness	A070909	Oversensing		
							E0752	Unspecified Respiratory Problem	A070909	Oversensing		
					CT983	Surgical invasive	E0112	Dizziness	A070909	Oversensing		
							E0752	Unspecified Respiratory Problem	A070909	Oversensing		
9412	2020	All other reportable incidents	35223	Endocardial/interventricular septal pacing lead	CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A040101	Fracture		
									A070909	Oversensing		
							F1901	Additional Surgery	A040101	Fracture		
									A070909	Oversensing		
							F1905	Device Revision or Replacement	A040101	Fracture		
									A070909	Oversensing		

9413	2020	All other reportable incidents	47263	Cardiac resynchronization therapy implantable pacemaker	CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070909	Oversensing		
							F26	No Health Consequences or Impact	A070909	Oversensing		
					CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A070909	Oversensing		
							F26	No Health Consequences or Impact	A070909	Oversensing		
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A070909	Oversensing		
							F26	No Health Consequences or Impact	A070909	Oversensing		
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A070909	Oversensing		
							F26	No Health Consequences or Impact	A070909	Oversensing		
9414	2020	All other reportable incidents	37265	Dual-chamber implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A072201	High Impedance		
							F24	Insufficient Health Impact Information	A072201	High Impedance		
9415	2020	All other reportable incidents	47265	Dual-chamber implantable pacemaker, rate-responsive	CT141	Electrodes/Leads and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	23/06/2020	23/06/2020
							F12	Serious Injury/ Illness/ Impairment	A070504	Premature Discharge of Battery	23/06/2020	23/06/2020
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	23/06/2020	23/06/2020
					CT145	Cardiovascular devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	23/06/2020	23/06/2020
							F12	Serious Injury/ Illness/ Impairment	A070504	Premature Discharge of Battery	23/06/2020	23/06/2020
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	23/06/2020	23/06/2020
					CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	23/06/2020	23/06/2020
							F12	Serious Injury/ Illness/ Impairment	A070504	Premature Discharge of Battery	23/06/2020	23/06/2020
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	23/06/2020	23/06/2020
					CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	23/06/2020	23/06/2020
							F12	Serious Injury/ Illness/ Impairment	A070504	Premature Discharge of Battery	23/06/2020	23/06/2020
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	23/06/2020	23/06/2020
					CT233	Surgical	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	23/06/2020	23/06/2020
							F12	Serious Injury/ Illness/ Impairment	A070504	Premature Discharge of Battery	23/06/2020	23/06/2020
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	23/06/2020	23/06/2020
					CT2406	Implantable	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	23/06/2020	23/06/2020
							F12	Serious Injury/ Illness/ Impairment	A070504	Premature Discharge of Battery	23/06/2020	23/06/2020
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	23/06/2020	23/06/2020
					CT2407	Active implantable	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	23/06/2020	23/06/2020
							F12	Serious Injury/ Illness/ Impairment	A070504	Premature Discharge of Battery	23/06/2020	23/06/2020
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	23/06/2020	23/06/2020
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	23/06/2020	23/06/2020
							F12	Serious Injury/ Illness/ Impairment	A070504	Premature Discharge of Battery	23/06/2020	23/06/2020
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	23/06/2020	23/06/2020
					CT2727	Active implant devices/ancillaries	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	23/06/2020	23/06/2020
							F12	Serious Injury/ Illness/ Impairment	A070504	Premature Discharge of Battery	23/06/2020	23/06/2020
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	23/06/2020	23/06/2020
					CT2750	Implantable cardiac electrical stimulators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	23/06/2020	23/06/2020
							F12	Serious Injury/ Illness/ Impairment	A070504	Premature Discharge of Battery	23/06/2020	23/06/2020
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	23/06/2020	23/06/2020
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	23/06/2020	23/06/2020
							F12	Serious Injury/ Illness/ Impairment	A070504	Premature Discharge of Battery	23/06/2020	23/06/2020
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	23/06/2020	23/06/2020
					CT2985	Active	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	23/06/2020	23/06/2020
							F12	Serious Injury/ Illness/ Impairment	A070504	Premature Discharge of Battery	23/06/2020	23/06/2020
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	23/06/2020	23/06/2020
					CT3001	Implantable metal	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	23/06/2020	23/06/2020
							F12	Serious Injury/ Illness/ Impairment	A070504	Premature Discharge of Battery	23/06/2020	23/06/2020
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	23/06/2020	23/06/2020
					CT3002	Implantable synthetic polymer	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	23/06/2020	23/06/2020
							F12	Serious Injury/ Illness/ Impairment	A070504	Premature Discharge of Battery	23/06/2020	23/06/2020
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	23/06/2020	23/06/2020
					CT3006	Implantable inorganic material	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	23/06/2020	23/06/2020
							F12	Serious Injury/ Illness/ Impairment	A070504	Premature Discharge of Battery	23/06/2020	23/06/2020
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	23/06/2020	23/06/2020
					CT314	Electrically powered	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	23/06/2020	23/06/2020
							F12	Serious Injury/ Illness/ Impairment	A070504	Premature Discharge of Battery	23/06/2020	23/06/2020
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	23/06/2020	23/06/2020
					CT321	Long-term surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	23/06/2020	23/06/2020
							F12	Serious Injury/ Illness/ Impairment	A070504	Premature Discharge of Battery	23/06/2020	23/06/2020
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	23/06/2020	23/06/2020
					CT449	Generators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	23/06/2020	23/06/2020
							F12	Serious Injury/ Illness/ Impairment	A070504	Premature Discharge of Battery	23/06/2020	23/06/2020
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	23/06/2020	23/06/2020
					CT981	Single use	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	23/06/2020	23/06/2020
							F12	Serious Injury/ Illness/ Impairment	A070504	Premature Discharge of Battery	23/06/2020	23/06/2020
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	23/06/2020	23/06/2020
					CT983	Surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	23/06/2020	23/06/2020
							F12	Serious Injury/ Illness/ Impairment	A070504	Premature Discharge of Battery	23/06/2020	23/06/2020
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	23/06/2020	23/06/2020
9416	2020	All other reportable incidents	17882	Physiologic-monitoring defibrillation system	CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070903	Failure to Analyze Signal		
							F02	Death	A070903	Failure to Analyze Signal		
9417	2020	All other reportable incidents	47270	Cardiac resynchronization therapy implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E2401	Insufficient Information	A160601	Premature Elective Replacement Indicator		
							F1905	Device Revision or Replacement	A160601	Premature Elective Replacement Indicator		
							F24	Insufficient Health Impact Information	A160601	Premature Elective Replacement Indicator		
9418	2020	All other reportable incidents	47267	Single-chamber implantable pacemaker, rate-responsive	CT141	Electrodes/Leads and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	21/12/2020	21/12/2020
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	21/12/2020	21/12/2020
							F12	Serious Injury/ Illness/ Impairment	A070504	Premature Discharge of Battery	21/12/2020	21/12/2020
							F1903	Device Explantation	A070504	Premature Discharge of Battery	21/12/2020	21/12/2020
					CT145	Cardiovascular devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	21/12/2020	21/12/2020
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	21/12/2020	21/12/2020
							F12	Serious Injury/ Illness/ Impairment	A070504	Premature Discharge of Battery	21/12/2020	21/12/2020
							F1903	Device Explantation	A070504	Premature Discharge of Battery	21/12/2020	21/12/2020
					CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	21/12/2020	21/12/2020
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	21/12/2020	21/12/2020
							F12	Serious Injury/ Illness/ Impairment	A070504	Premature Discharge of Battery	21/12/2020	21/12/2020
							F1903	Device Explantation	A070504	Premature Discharge of Battery	21/12/2020	21/12/2020
					CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	21/12/2020	21/12/2020

							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	21/12/2020	21/12/2020
							F12	Serious Injury/ Illness/ Impairment	A070504	Premature Discharge of Battery	21/12/2020	21/12/2020
							F1903	Device Explantation	A070504	Premature Discharge of Battery	21/12/2020	21/12/2020
					CT233	Surgical	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	21/12/2020	21/12/2020
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	21/12/2020	21/12/2020
							F12	Serious Injury/ Illness/ Impairment	A070504	Premature Discharge of Battery	21/12/2020	21/12/2020
							F1903	Device Explantation	A070504	Premature Discharge of Battery	21/12/2020	21/12/2020
					CT2406	Implantable	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	21/12/2020	21/12/2020
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	21/12/2020	21/12/2020
							F12	Serious Injury/ Illness/ Impairment	A070504	Premature Discharge of Battery	21/12/2020	21/12/2020
							F1903	Device Explantation	A070504	Premature Discharge of Battery	21/12/2020	21/12/2020
					CT2407	Active implantable	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	21/12/2020	21/12/2020
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	21/12/2020	21/12/2020
							F12	Serious Injury/ Illness/ Impairment	A070504	Premature Discharge of Battery	21/12/2020	21/12/2020
							F1903	Device Explantation	A070504	Premature Discharge of Battery	21/12/2020	21/12/2020
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	21/12/2020	21/12/2020
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	21/12/2020	21/12/2020
							F12	Serious Injury/ Illness/ Impairment	A070504	Premature Discharge of Battery	21/12/2020	21/12/2020
							F1903	Device Explantation	A070504	Premature Discharge of Battery	21/12/2020	21/12/2020
					CT2727	Active implant devices/ancillaries	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	21/12/2020	21/12/2020
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	21/12/2020	21/12/2020
							F12	Serious Injury/ Illness/ Impairment	A070504	Premature Discharge of Battery	21/12/2020	21/12/2020
							F1903	Device Explantation	A070504	Premature Discharge of Battery	21/12/2020	21/12/2020
					CT2750	Implantable cardiac electrical stimulators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	21/12/2020	21/12/2020
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	21/12/2020	21/12/2020
							F12	Serious Injury/ Illness/ Impairment	A070504	Premature Discharge of Battery	21/12/2020	21/12/2020
							F1903	Device Explantation	A070504	Premature Discharge of Battery	21/12/2020	21/12/2020
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	21/12/2020	21/12/2020
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	21/12/2020	21/12/2020
							F12	Serious Injury/ Illness/ Impairment	A070504	Premature Discharge of Battery	21/12/2020	21/12/2020
							F1903	Device Explantation	A070504	Premature Discharge of Battery	21/12/2020	21/12/2020
					CT2985	Active	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	21/12/2020	21/12/2020
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	21/12/2020	21/12/2020
							F12	Serious Injury/ Illness/ Impairment	A070504	Premature Discharge of Battery	21/12/2020	21/12/2020
							F1903	Device Explantation	A070504	Premature Discharge of Battery	21/12/2020	21/12/2020
					CT3001	Implantable metal	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	21/12/2020	21/12/2020
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	21/12/2020	21/12/2020
							F12	Serious Injury/ Illness/ Impairment	A070504	Premature Discharge of Battery	21/12/2020	21/12/2020
							F1903	Device Explantation	A070504	Premature Discharge of Battery	21/12/2020	21/12/2020
					CT3002	Implantable synthetic polymer	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	21/12/2020	21/12/2020
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	21/12/2020	21/12/2020
							F12	Serious Injury/ Illness/ Impairment	A070504	Premature Discharge of Battery	21/12/2020	21/12/2020
							F1903	Device Explantation	A070504	Premature Discharge of Battery	21/12/2020	21/12/2020
					CT3006	Implantable inorganic material	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	21/12/2020	21/12/2020
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	21/12/2020	21/12/2020
							F12	Serious Injury/ Illness/ Impairment	A070504	Premature Discharge of Battery	21/12/2020	21/12/2020
							F1903	Device Explantation	A070504	Premature Discharge of Battery	21/12/2020	21/12/2020
					CT314	Electrically powered	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	21/12/2020	21/12/2020
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	21/12/2020	21/12/2020
							F12	Serious Injury/ Illness/ Impairment	A070504	Premature Discharge of Battery	21/12/2020	21/12/2020
							F1903	Device Explantation	A070504	Premature Discharge of Battery	21/12/2020	21/12/2020
					CT321	Long-term surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	21/12/2020	21/12/2020
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	21/12/2020	21/12/2020
							F12	Serious Injury/ Illness/ Impairment	A070504	Premature Discharge of Battery	21/12/2020	21/12/2020
							F1903	Device Explantation	A070504	Premature Discharge of Battery	21/12/2020	21/12/2020
					CT449	Generators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	21/12/2020	21/12/2020
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	21/12/2020	21/12/2020
							F12	Serious Injury/ Illness/ Impairment	A070504	Premature Discharge of Battery	21/12/2020	21/12/2020
							F1903	Device Explantation	A070504	Premature Discharge of Battery	21/12/2020	21/12/2020
					CT981	Single use	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	21/12/2020	21/12/2020
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	21/12/2020	21/12/2020
							F12	Serious Injury/ Illness/ Impairment	A070504	Premature Discharge of Battery	21/12/2020	21/12/2020
							F1903	Device Explantation	A070504	Premature Discharge of Battery	21/12/2020	21/12/2020
					CT983	Surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	21/12/2020	21/12/2020
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	21/12/2020	21/12/2020
							F12	Serious Injury/ Illness/ Impairment	A070504	Premature Discharge of Battery	21/12/2020	21/12/2020
							F1903	Device Explantation	A070504	Premature Discharge of Battery	21/12/2020	21/12/2020
9419	2019	All other reportable incidents	37265	Dual-chamber implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	02/08/2019	
9420	2019	All other reportable incidents	35223	Endocardial/interventricular septal pacing lead	CT1684	Cardiac pacemakers and associated devices	F1905	Device Revision or Replacement	A0701	Capturing Problem		
9421	2019	All other reportable incidents	37265	Dual-chamber implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A070909	Oversensing	16/02/2021	16/02/2021
									A071302	Inappropriate/Inadequate Shock/Stimulation	16/02/2021	16/02/2021
							F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A070909	Oversensing	16/02/2021	16/02/2021
									A071302	Inappropriate/Inadequate Shock/Stimulation	16/02/2021	16/02/2021
							F26	No Health Consequences or Impact	A070909	Oversensing	16/02/2021	16/02/2021
									A071302	Inappropriate/Inadequate Shock/Stimulation	16/02/2021	16/02/2021
9422	2020	All other reportable incidents	35223	Endocardial/interventricular septal pacing lead	CT1684	Cardiac pacemakers and associated devices	E060109	Tachycardia	A070901	Decreased Sensitivity		
									A0722	Impedance Problem		
									A072201	High Impedance		
									A090801	Signal Artifact/Noise		
							E2403	No Clinical Signs, Symptoms or Conditions	A070901	Decreased Sensitivity		
									A0722	Impedance Problem		
									A072201	High Impedance		
									A090801	Signal Artifact/Noise		
							F26	No Health Consequences or Impact	A070901	Decreased Sensitivity		
									A0722	Impedance Problem		
									A072201	High Impedance		
									A090801	Signal Artifact/Noise		

9423	2020	All other reportable incidents	47270	Cardiac resynchronization therapy implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery		
							F1903	Device Explantation	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
9424	2020	All other reportable incidents	17882	Physiologic-monitoring defibrillation system	CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A0719	Unexpected Shutdown		
							F26	No Health Consequences or Impact	A0719	Unexpected Shutdown		
9425	2020	All other reportable incidents	47270	Cardiac resynchronization therapy implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E0601	Arrhythmia	A25	No Apparent Adverse Event		
							F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A25	No Apparent Adverse Event		
							F12	Serious Injury/ Illness/ Impairment	A25	No Apparent Adverse Event		
							F14	Prolonged Episode of Care	A25	No Apparent Adverse Event		
9426	2020	All other reportable incidents	47265	Dual-chamber implantable pacemaker, rate-responsive	CT141	Electrodes/Leads and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT145	Cardiovascular devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT233	Surgical	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT2406	Implantable	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT2407	Active implantable	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT2727	Active implant devices/ancillaries	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT2750	Implantable cardiac electrical stimulators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT2985	Active	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT3001	Implantable metal	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT3002	Implantable synthetic polymer	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT3006	Implantable inorganic material	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT314	Electrically powered	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT321	Long-term surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT449	Generators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT981	Single use	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT983	Surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
9427	2020	All other reportable incidents	35852	Single-chamber implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A070909	Oversensing		
							F26	No Health Consequences or Impact	A070909	Oversensing		
9428	2020	All other reportable incidents	35852	Single-chamber implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E2104	Electric Shock	A071302	Inappropriate/Inadequate Shock/Stimulation		
							E2403	No Clinical Signs, Symptoms or Conditions	A071302	Inappropriate/Inadequate Shock/Stimulation		
							F26	No Health Consequences or Impact	A071302	Inappropriate/Inadequate Shock/Stimulation		
9429	2020	All other reportable incidents	36241	Pacemaker/defibrillator lead	CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070909	Oversensing		
							F08	Hospitalization or Prolonged Hospitalization	A070909	Oversensing		
							F15	Recognized Device or Procedural Complication	A070909	Oversensing		
9430	2020	All other reportable incidents	36106	Implantable pacemaker analyser, intraoperative	CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A0709	Device Sensing Problem		
									A070908	Failure to Sense		
							F1905	Device Revision or Replacement	A0709	Device Sensing Problem		
									A070908	Failure to Sense		
							F26	No Health Consequences or Impact	A0709	Device Sensing Problem		
									A070908	Failure to Sense		
9431	2019	All other reportable incidents	47267	Single-chamber implantable pacemaker, rate-responsive	CT141	Electrodes/Leads and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
					CT145	Cardiovascular devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
					CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
					CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
					CT233	Surgical	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
					CT2406	Implantable	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
					CT2407	Active implantable	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
					CT2727	Active implant devices/ancillaries	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
					CT2750	Implantable cardiac electrical stimulators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
					CT2985	Active	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
					CT3001	Implantable metal	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
					CT3002	Implantable synthetic polymer	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
					CT3006	Implantable inorganic material	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
					CT314	Electrically powered	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
					CT321	Long-term surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
					CT449	Generators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
					CT981	Single use	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
					CT983	Surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
9432	2019	All other reportable incidents	36241	Pacemaker/defibrillator lead	CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A0510	Retraction Problem		
9433	2019	All other reportable incidents	35852	Single-chamber implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A072201	High Impedance		
9434	2019	All other reportable incidents	35852	Single-chamber implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A070909	Oversensing		

9435	2020	All other reportable incidents	35852	Single-chamber implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F1901	Additional Surgery	A070504	Premature Discharge of Battery		
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery		
9436	2020	All other reportable incidents	47270	Cardiac resynchronization therapy implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	26/06/2020	26/06/2020
							F1901	Additional Surgery	A070504	Premature Discharge of Battery	26/06/2020	26/06/2020
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	26/06/2020	26/06/2020
9437	2019	All other reportable incidents	47265	Dual-chamber implantable pacemaker, rate-responsive	CT141	Electrodes/Leads and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070909	Oversensing		
					CT145	Cardiovascular devices	E2403	No Clinical Signs, Symptoms or Conditions	A070909	Oversensing		
					CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070909	Oversensing		
					CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A070909	Oversensing		
					CT233	Surgical	E2403	No Clinical Signs, Symptoms or Conditions	A070909	Oversensing		
					CT2406	Implantable	E2403	No Clinical Signs, Symptoms or Conditions	A070909	Oversensing		
					CT2407	Active implantable	E2403	No Clinical Signs, Symptoms or Conditions	A070909	Oversensing		
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A070909	Oversensing		
					CT2727	Active implant devices/ancillaries	E2403	No Clinical Signs, Symptoms or Conditions	A070909	Oversensing		
					CT2750	Implantable cardiac electrical stimulators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070909	Oversensing		
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A070909	Oversensing		
					CT2985	Active	E2403	No Clinical Signs, Symptoms or Conditions	A070909	Oversensing		
					CT3001	Implantable metal	E2403	No Clinical Signs, Symptoms or Conditions	A070909	Oversensing		
					CT3002	Implantable synthetic polymer	E2403	No Clinical Signs, Symptoms or Conditions	A070909	Oversensing		
					CT3006	Implantable inorganic material	E2403	No Clinical Signs, Symptoms or Conditions	A070909	Oversensing		
					CT314	Electrically powered	E2403	No Clinical Signs, Symptoms or Conditions	A070909	Oversensing		
					CT321	Long-term surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A070909	Oversensing		
					CT449	Generators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070909	Oversensing		
					CT981	Single use	E2403	No Clinical Signs, Symptoms or Conditions	A070909	Oversensing		
					CT983	Surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A070909	Oversensing		
9438	2019	All other reportable incidents	37265	Dual-chamber implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A071302	Inappropriate/Inadequate Shock/Stimulation		
9439	2019	All other reportable incidents	35852	Single-chamber implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A070909	Oversensing		
9440	2019	All other reportable incidents	35223	Endocardial/interventricular septal pacing lead	CT1684	Cardiac pacemakers and associated devices	F1905	Device Revision or Replacement	A070909	Oversensing		
9441	2020	All other reportable incidents	35852	Single-chamber implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	26/06/2020	26/06/2020
									A090805	Incorrect Measurement	26/06/2020	26/06/2020
							F1901	Additional Surgery	A070504	Premature Discharge of Battery	26/06/2020	26/06/2020
									A090805	Incorrect Measurement	26/06/2020	26/06/2020
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	26/06/2020	26/06/2020
									A090805	Incorrect Measurement	26/06/2020	26/06/2020
9442	2020	All other reportable incidents	35852	Single-chamber implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A070602	Delayed Charge Time		
							F1905	Device Revision or Replacement	A070602	Delayed Charge Time		
							F26	No Health Consequences or Impact	A070602	Delayed Charge Time		
9443	2019	All other reportable incidents	47270	Cardiac resynchronization therapy implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	24/07/2019	
9444	2019	All other reportable incidents	37265	Dual-chamber implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A0709	Device Sensing Problem		
9445	2019	All other reportable incidents	17882	Physiologic-monitoring defibrillation system	CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070801	Complete Loss of Power		
9446	2020	All other reportable incidents	47270	Cardiac resynchronization therapy implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A070910	Undersensing		
							F26	No Health Consequences or Impact	A070910	Undersensing		
9447	2019	All other reportable incidents	17882	Physiologic-monitoring defibrillation system	CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070908	Failure to Sense		
							F26	No Health Consequences or Impact	A070908	Failure to Sense		
9448	2019	All other reportable incidents	36241	Pacemaker/defibrillator lead	CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A0510	Retraction Problem		
9449	2020	All other reportable incidents	37265	Dual-chamber implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A072201	High Impedance		
							F26	No Health Consequences or Impact	A072201	High Impedance		
9450	2020	All other reportable incidents	47270	Cardiac resynchronization therapy implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A072201	High Impedance		
							F26	No Health Consequences or Impact	A072201	High Impedance		
9451	2019	All other reportable incidents	37265	Dual-chamber implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	26/07/2019	
9452	2020	All other reportable incidents	36241	Pacemaker/defibrillator lead	CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A072201	High Impedance	11/06/2020	11/06/2020
							F1905	Device Revision or Replacement	A072201	High Impedance	11/06/2020	11/06/2020
							F26	No Health Consequences or Impact	A072201	High Impedance	11/06/2020	11/06/2020
9453	2019	All other reportable incidents	47270	Cardiac resynchronization therapy implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A0722	Impedance Problem		
9454	2019	All other reportable incidents	35852	Single-chamber implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	08/08/2019	
9455	2020	All other reportable incidents	37265	Dual-chamber implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A070910	Undersensing		
							F26	No Health Consequences or Impact	A070910	Undersensing		
9456	2019	All other reportable incidents	47267	Single-chamber implantable pacemaker, rate-responsive	CT141	Electrodes/Leads and associated devices	F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery		
					CT145	Cardiovascular devices	F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery		
					CT1684	Cardiac pacemakers and associated devices	F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery		
					CT1685	Cardiac pacemakers	F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery		
					CT233	Surgical	F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery		
					CT2406	Implantable	F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery		
					CT2407	Active implantable	F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery		
					CT257	Implantable cardiac pacemakers	F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery		
					CT2727	Active implant devices/ancillaries	F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery		
					CT2750	Implantable cardiac electrical stimulators and associated devices	F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery		
					CT2866	Implantable cardiac electrical stimulators	F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery		
					CT2985	Active	F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery		
					CT3001	Implantable metal	F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery		
					CT3002	Implantable synthetic polymer	F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery		
					CT3006	Implantable inorganic material	F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery		
					CT314	Electrically powered	F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery		
					CT321	Long-term surgical invasive	F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery		
					CT449	Generators and associated devices	F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery		
					CT981	Single use	F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery		
					CT983	Surgical invasive	F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery		
9457	2019	All other reportable incidents	37265	Dual-chamber implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	F1905	Device Revision or Replacement	A070603	Failure to Charge	30/07/2019	
9458	2019	All other reportable incidents	47263	Cardiac resynchronization therapy implantable pacemaker	CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	12/07/2019	12/07/2019
					CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	12/07/2019	12/07/2019
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	12/07/2019	12/07/2019
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	12/07/2019	12/07/2019
9459	2020	All other reportable incidents	61438	Cardiac pulse generator reader, home-use	CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A1005	Overheating of Device		

							F26	No Health Consequences or Impact	A1005	Overheating of Device		
9460	2019	All other reportable incidents	37265	Dual-chamber implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	07/08/2019	
9461	2020	All other reportable incidents	35852	Single-chamber implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A071302	Inappropriate/Inadequate Shock/Stimulation		
							F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A071302	Inappropriate/Inadequate Shock/Stimulation		
							F26	No Health Consequences or Impact	A071302	Inappropriate/Inadequate Shock/Stimulation		
9462	2020	All other reportable incidents	37265	Dual-chamber implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A072201	High Impedance		
							F26	No Health Consequences or Impact	A072201	High Impedance		
9463	2019	All other reportable incidents	37265	Dual-chamber implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A070910	Undersensing		
9464	2019	All other reportable incidents	35852	Single-chamber implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A070909	Oversensing		
9465	2019	All other reportable incidents	47265	Dual-chamber implantable pacemaker, rate-responsive	CT141	Electrodes/Leads and associated devices	F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	01/08/2019	
					CT145	Cardiovascular devices	F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	01/08/2019	
					CT1684	Cardiac pacemakers and associated devices	F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	01/08/2019	
					CT1685	Cardiac pacemakers	F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	01/08/2019	
					CT233	Surgical	F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	01/08/2019	
					CT2406	Implantable	F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	01/08/2019	
					CT2407	Active implantable	F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	01/08/2019	
					CT257	Implantable cardiac pacemakers	F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	01/08/2019	
					CT2727	Active implant devices/ancillaries	F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	01/08/2019	
					CT2750	Implantable cardiac electrical stimulators and associated devices	F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	01/08/2019	
					CT2866	Implantable cardiac electrical stimulators	F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	01/08/2019	
					CT2985	Active	F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	01/08/2019	
					CT3001	Implantable metal	F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	01/08/2019	
					CT3002	Implantable synthetic polymer	F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	01/08/2019	
					CT3006	Implantable inorganic material	F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	01/08/2019	
					CT314	Electrically powered	F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	01/08/2019	
					CT321	Long-lerm surgical invasive	F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	01/08/2019	
					CT449	Generators and associated devices	F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	01/08/2019	
					CT981	Single use	F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	01/08/2019	
					CT983	Surgical invasive	F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	01/08/2019	
9466	2020	All other reportable incidents	17882	Physiologic-monitoring defibrillation system	CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070803	Failure to Power Up		
							F27	Problem Identified During Non-Clinical Procedure	A070803	Failure to Power Up		
9467	2019	All other reportable incidents	35852	Single-chamber implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A070909	Oversensing		
9468	2019	All other reportable incidents	35852	Single-chamber implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A071302	Inappropriate/Inadequate Shock/Stimulation		
9469	2019	All other reportable incidents	36241	Pacemaker/defibrillator lead	CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070101	Failure to Capture		
9470	2019	All other reportable incidents	17882	Physiologic-monitoring defibrillation system	CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070908	Failure to Sense		
							F26	No Health Consequences or Impact	A070908	Failure to Sense		
9471	2019	All other reportable incidents	36241	Pacemaker/defibrillator lead	CT1684	Cardiac pacemakers and associated devices	F1905	Device Revision or Replacement	A072201	High Impedance		
9472	2020	All other reportable incidents	47265	Dual-chamber implantable pacemaker, rate-responsive	CT141	Electrodes/Leads and associated devices	E0112	Dizziness	A27	Appropriate Device Problem Term/Code not Available		
							E0204	Lethargy	A27	Appropriate Device Problem Term/Code not Available		
							E0717	Dyspnea	A27	Appropriate Device Problem Term/Code not Available		
							E0752	Unspecified Respiratory Problem	A27	Appropriate Device Problem Term/Code not Available		
							E1032	Vomiting	A27	Appropriate Device Problem Term/Code not Available		
							E2312	Fatigue	A27	Appropriate Device Problem Term/Code not Available		
							E233001	Chest Pain	A27	Appropriate Device Problem Term/Code not Available		
							F11	Non-Serious Injury/ Illness/ Impairment	A27	Appropriate Device Problem Term/Code not Available		
					CT145	Cardiovascular devices	E0112	Dizziness	A27	Appropriate Device Problem Term/Code not Available		
							E0204	Lethargy	A27	Appropriate Device Problem Term/Code not Available		
							E0717	Dyspnea	A27	Appropriate Device Problem Term/Code not Available		
							E0752	Unspecified Respiratory Problem	A27	Appropriate Device Problem Term/Code not Available		
							E1032	Vomiting	A27	Appropriate Device Problem Term/Code not Available		
							E2312	Fatigue	A27	Appropriate Device Problem Term/Code not Available		
							E233001	Chest Pain	A27	Appropriate Device Problem Term/Code not Available		
							F11	Non-Serious Injury/ Illness/ Impairment	A27	Appropriate Device Problem Term/Code not Available		
					CT1684	Cardiac pacemakers and associated devices	E0112	Dizziness	A27	Appropriate Device Problem Term/Code not Available		
							E0204	Lethargy	A27	Appropriate Device Problem Term/Code not Available		
							E0717	Dyspnea	A27	Appropriate Device Problem Term/Code not Available		
							E0752	Unspecified Respiratory Problem	A27	Appropriate Device Problem Term/Code not Available		
							E1032	Vomiting	A27	Appropriate Device Problem Term/Code not Available		
							E2312	Fatigue	A27	Appropriate Device Problem Term/Code not Available		
							E233001	Chest Pain	A27	Appropriate Device Problem Term/Code not Available		
							F11	Non-Serious Injury/ Illness/ Impairment	A27	Appropriate Device Problem Term/Code not Available		
					CT1685	Cardiac pacemakers	E0112	Dizziness	A27	Appropriate Device Problem Term/Code not Available		
							E0204	Lethargy	A27	Appropriate Device Problem Term/Code not Available		
							E0717	Dyspnea	A27	Appropriate Device Problem Term/Code not Available		
							E0752	Unspecified Respiratory Problem	A27	Appropriate Device Problem Term/Code not Available		

							E1032	Vomiting	A27	Appropriate Device Problem Term/Code not Available		
							E2312	Fatigue	A27	Appropriate Device Problem Term/Code not Available		
							E233001	Chest Pain	A27	Appropriate Device Problem Term/Code not Available		
							F11	Non-Serious Injury/ Illness/ Impairment	A27	Appropriate Device Problem Term/Code not Available		
					CT233	Surgical	E0112	Dizziness	A27	Appropriate Device Problem Term/Code not Available		
							E0204	Lethargy	A27	Appropriate Device Problem Term/Code not Available		
							E0717	Dyspnea	A27	Appropriate Device Problem Term/Code not Available		
							E0752	Unspecified Respiratory Problem	A27	Appropriate Device Problem Term/Code not Available		
							E1032	Vomiting	A27	Appropriate Device Problem Term/Code not Available		
							E2312	Fatigue	A27	Appropriate Device Problem Term/Code not Available		
							E233001	Chest Pain	A27	Appropriate Device Problem Term/Code not Available		
							F11	Non-Serious Injury/ Illness/ Impairment	A27	Appropriate Device Problem Term/Code not Available		
					CT2406	Implantable	E0112	Dizziness	A27	Appropriate Device Problem Term/Code not Available		
							E0204	Lethargy	A27	Appropriate Device Problem Term/Code not Available		
							E0717	Dyspnea	A27	Appropriate Device Problem Term/Code not Available		
							E0752	Unspecified Respiratory Problem	A27	Appropriate Device Problem Term/Code not Available		
							E1032	Vomiting	A27	Appropriate Device Problem Term/Code not Available		
							E2312	Fatigue	A27	Appropriate Device Problem Term/Code not Available		
							E233001	Chest Pain	A27	Appropriate Device Problem Term/Code not Available		
							F11	Non-Serious Injury/ Illness/ Impairment	A27	Appropriate Device Problem Term/Code not Available		
					CT2407	Active implantable	E0112	Dizziness	A27	Appropriate Device Problem Term/Code not Available		
							E0204	Lethargy	A27	Appropriate Device Problem Term/Code not Available		
							E0717	Dyspnea	A27	Appropriate Device Problem Term/Code not Available		
							E0752	Unspecified Respiratory Problem	A27	Appropriate Device Problem Term/Code not Available		
							E1032	Vomiting	A27	Appropriate Device Problem Term/Code not Available		
							E2312	Fatigue	A27	Appropriate Device Problem Term/Code not Available		
							E233001	Chest Pain	A27	Appropriate Device Problem Term/Code not Available		
							F11	Non-Serious Injury/ Illness/ Impairment	A27	Appropriate Device Problem Term/Code not Available		
					CT257	Implantable cardiac pacemakers	E0112	Dizziness	A27	Appropriate Device Problem Term/Code not Available		
							E0204	Lethargy	A27	Appropriate Device Problem Term/Code not Available		
							E0717	Dyspnea	A27	Appropriate Device Problem Term/Code not Available		
							E0752	Unspecified Respiratory Problem	A27	Appropriate Device Problem Term/Code not Available		
							E1032	Vomiting	A27	Appropriate Device Problem Term/Code not Available		
							E2312	Fatigue	A27	Appropriate Device Problem Term/Code not Available		
							E233001	Chest Pain	A27	Appropriate Device Problem Term/Code not Available		
							F11	Non-Serious Injury/ Illness/ Impairment	A27	Appropriate Device Problem Term/Code not Available		
					CT2727	Active implant devices/ancillaries	E0112	Dizziness	A27	Appropriate Device Problem Term/Code not Available		
							E0204	Lethargy	A27	Appropriate Device Problem Term/Code not Available		
							E0717	Dyspnea	A27	Appropriate Device Problem Term/Code not Available		
							E0752	Unspecified Respiratory Problem	A27	Appropriate Device Problem Term/Code not Available		
							E1032	Vomiting	A27	Appropriate Device Problem Term/Code not Available		
							E2312	Fatigue	A27	Appropriate Device Problem Term/Code not Available		
							E233001	Chest Pain	A27	Appropriate Device Problem Term/Code not Available		
							F11	Non-Serious Injury/ Illness/ Impairment	A27	Appropriate Device Problem Term/Code not Available		
					CT2750	Implantable cardiac electrical stimulators and associated devices	E0112	Dizziness	A27	Appropriate Device Problem Term/Code not Available		
							E0204	Lethargy	A27	Appropriate Device Problem Term/Code not Available		
							E0717	Dyspnea	A27	Appropriate Device Problem Term/Code not Available		
							E0752	Unspecified Respiratory Problem	A27	Appropriate Device Problem Term/Code not Available		
							E1032	Vomiting	A27	Appropriate Device Problem Term/Code not Available		
							E2312	Fatigue	A27	Appropriate Device Problem Term/Code not Available		

							E233001	Chest Pain	A27	Appropriate Device Problem Term/Code not Available		
							F11	Non-Serious Injury/ Illness/ Impairment	A27	Appropriate Device Problem Term/Code not Available		
					CT2866	Implantable cardiac electrical stimulators	E0112	Dizziness	A27	Appropriate Device Problem Term/Code not Available		
							E0204	Lethargy	A27	Appropriate Device Problem Term/Code not Available		
							E0717	Dyspnea	A27	Appropriate Device Problem Term/Code not Available		
							E0752	Unspecified Respiratory Problem	A27	Appropriate Device Problem Term/Code not Available		
							E1032	Vomiting	A27	Appropriate Device Problem Term/Code not Available		
							E2312	Fatigue	A27	Appropriate Device Problem Term/Code not Available		
							E233001	Chest Pain	A27	Appropriate Device Problem Term/Code not Available		
							F11	Non-Serious Injury/ Illness/ Impairment	A27	Appropriate Device Problem Term/Code not Available		
					CT2985	Active	E0112	Dizziness	A27	Appropriate Device Problem Term/Code not Available		
							E0204	Lethargy	A27	Appropriate Device Problem Term/Code not Available		
							E0717	Dyspnea	A27	Appropriate Device Problem Term/Code not Available		
							E0752	Unspecified Respiratory Problem	A27	Appropriate Device Problem Term/Code not Available		
							E1032	Vomiting	A27	Appropriate Device Problem Term/Code not Available		
							E2312	Fatigue	A27	Appropriate Device Problem Term/Code not Available		
							E233001	Chest Pain	A27	Appropriate Device Problem Term/Code not Available		
							F11	Non-Serious Injury/ Illness/ Impairment	A27	Appropriate Device Problem Term/Code not Available		
					CT3001	Implantable metal	E0112	Dizziness	A27	Appropriate Device Problem Term/Code not Available		
							E0204	Lethargy	A27	Appropriate Device Problem Term/Code not Available		
							E0717	Dyspnea	A27	Appropriate Device Problem Term/Code not Available		
							E0752	Unspecified Respiratory Problem	A27	Appropriate Device Problem Term/Code not Available		
							E1032	Vomiting	A27	Appropriate Device Problem Term/Code not Available		
							E2312	Fatigue	A27	Appropriate Device Problem Term/Code not Available		
							E233001	Chest Pain	A27	Appropriate Device Problem Term/Code not Available		
							F11	Non-Serious Injury/ Illness/ Impairment	A27	Appropriate Device Problem Term/Code not Available		
					CT3002	Implantable synthetic polymer	E0112	Dizziness	A27	Appropriate Device Problem Term/Code not Available		
							E0204	Lethargy	A27	Appropriate Device Problem Term/Code not Available		
							E0717	Dyspnea	A27	Appropriate Device Problem Term/Code not Available		
							E0752	Unspecified Respiratory Problem	A27	Appropriate Device Problem Term/Code not Available		
							E1032	Vomiting	A27	Appropriate Device Problem Term/Code not Available		
							E2312	Fatigue	A27	Appropriate Device Problem Term/Code not Available		
							E233001	Chest Pain	A27	Appropriate Device Problem Term/Code not Available		
							F11	Non-Serious Injury/ Illness/ Impairment	A27	Appropriate Device Problem Term/Code not Available		
					CT3006	Implantable inorganic material	E0112	Dizziness	A27	Appropriate Device Problem Term/Code not Available		
							E0204	Lethargy	A27	Appropriate Device Problem Term/Code not Available		
							E0717	Dyspnea	A27	Appropriate Device Problem Term/Code not Available		
							E0752	Unspecified Respiratory Problem	A27	Appropriate Device Problem Term/Code not Available		
							E1032	Vomiting	A27	Appropriate Device Problem Term/Code not Available		
							E2312	Fatigue	A27	Appropriate Device Problem Term/Code not Available		
							E233001	Chest Pain	A27	Appropriate Device Problem Term/Code not Available		
							F11	Non-Serious Injury/ Illness/ Impairment	A27	Appropriate Device Problem Term/Code not Available		
					CT314	Electrically powered	E0112	Dizziness	A27	Appropriate Device Problem Term/Code not Available		
							E0204	Lethargy	A27	Appropriate Device Problem Term/Code not Available		
							E0717	Dyspnea	A27	Appropriate Device Problem Term/Code not Available		
							E0752	Unspecified Respiratory Problem	A27	Appropriate Device Problem Term/Code not Available		
							E1032	Vomiting	A27	Appropriate Device Problem Term/Code not Available		
							E2312	Fatigue	A27	Appropriate Device Problem Term/Code not Available		
							E233001	Chest Pain	A27	Appropriate Device Problem Term/Code not Available		
							F11	Non-Serious Injury/ Illness/ Impairment	A27	Appropriate Device Problem Term/Code not Available		

					CT321	Long-term surgical invasive	E0112	Dizziness	A27	Appropriate Device Problem Term/Code not Available		
							E0204	Lethargy	A27	Appropriate Device Problem Term/Code not Available		
							E0717	Dyspnea	A27	Appropriate Device Problem Term/Code not Available		
							E0752	Unspecified Respiratory Problem	A27	Appropriate Device Problem Term/Code not Available		
							E1032	Vomiting	A27	Appropriate Device Problem Term/Code not Available		
							E2312	Fatigue	A27	Appropriate Device Problem Term/Code not Available		
							E233001	Chest Pain	A27	Appropriate Device Problem Term/Code not Available		
							F11	Non-Serious Injury/ Illness/ Impairment	A27	Appropriate Device Problem Term/Code not Available		
					CT449	Generators and associated devices	E0112	Dizziness	A27	Appropriate Device Problem Term/Code not Available		
							E0204	Lethargy	A27	Appropriate Device Problem Term/Code not Available		
							E0717	Dyspnea	A27	Appropriate Device Problem Term/Code not Available		
							E0752	Unspecified Respiratory Problem	A27	Appropriate Device Problem Term/Code not Available		
							E1032	Vomiting	A27	Appropriate Device Problem Term/Code not Available		
							E2312	Fatigue	A27	Appropriate Device Problem Term/Code not Available		
							E233001	Chest Pain	A27	Appropriate Device Problem Term/Code not Available		
							F11	Non-Serious Injury/ Illness/ Impairment	A27	Appropriate Device Problem Term/Code not Available		
					CT981	Single use	E0112	Dizziness	A27	Appropriate Device Problem Term/Code not Available		
							E0204	Lethargy	A27	Appropriate Device Problem Term/Code not Available		
							E0717	Dyspnea	A27	Appropriate Device Problem Term/Code not Available		
							E0752	Unspecified Respiratory Problem	A27	Appropriate Device Problem Term/Code not Available		
							E1032	Vomiting	A27	Appropriate Device Problem Term/Code not Available		
							E2312	Fatigue	A27	Appropriate Device Problem Term/Code not Available		
							E233001	Chest Pain	A27	Appropriate Device Problem Term/Code not Available		
							F11	Non-Serious Injury/ Illness/ Impairment	A27	Appropriate Device Problem Term/Code not Available		
					CT983	Surgical invasive	E0112	Dizziness	A27	Appropriate Device Problem Term/Code not Available		
							E0204	Lethargy	A27	Appropriate Device Problem Term/Code not Available		
							E0717	Dyspnea	A27	Appropriate Device Problem Term/Code not Available		
							E0752	Unspecified Respiratory Problem	A27	Appropriate Device Problem Term/Code not Available		
							E1032	Vomiting	A27	Appropriate Device Problem Term/Code not Available		
							E2312	Fatigue	A27	Appropriate Device Problem Term/Code not Available		
							E233001	Chest Pain	A27	Appropriate Device Problem Term/Code not Available		
							F11	Non-Serious Injury/ Illness/ Impairment	A27	Appropriate Device Problem Term/Code not Available		
9473	2019	All other reportable incidents	35852	Single-chamber implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	03/08/2019	
9474	2020	All other reportable incidents	47265	Dual-chamber implantable pacemaker, rate-responsive	CT141	Electrodes/Leads and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
									A070909	Oversensing		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
									A070909	Oversensing		
					CT145	Cardiovascular devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
									A070909	Oversensing		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
									A070909	Oversensing		
					CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
									A070909	Oversensing		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
									A070909	Oversensing		
					CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
									A070909	Oversensing		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
									A070909	Oversensing		
					CT233	Surgical	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
									A070909	Oversensing		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
									A070909	Oversensing		
					CT2406	Implantable	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
									A070909	Oversensing		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
									A070909	Oversensing		
					CT2407	Active implantable	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
									A070909	Oversensing		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
									A070909	Oversensing		
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
									A070909	Oversensing		

							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
									A070909	Oversensing		
					CT2727	Active implant devices/ancillaries	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
									A070909	Oversensing		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
									A070909	Oversensing		
					CT2750	Implantable cardiac electrical stimulators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
									A070909	Oversensing		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
									A070909	Oversensing		
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
									A070909	Oversensing		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
									A070909	Oversensing		
					CT2985	Active	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
									A070909	Oversensing		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
									A070909	Oversensing		
					CT3001	Implantable metal	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
									A070909	Oversensing		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
									A070909	Oversensing		
					CT3002	Implantable synthetic polymer	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
									A070909	Oversensing		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
									A070909	Oversensing		
					CT3006	Implantable inorganic material	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
									A070909	Oversensing		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
									A070909	Oversensing		
					CT314	Electrically powered	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
									A070909	Oversensing		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
									A070909	Oversensing		
					CT321	Long-term surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
									A070909	Oversensing		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
									A070909	Oversensing		
					CT449	Generators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
									A070909	Oversensing		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
									A070909	Oversensing		
					CT981	Single use	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
									A070909	Oversensing		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
									A070909	Oversensing		
					CT983	Surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
									A070909	Oversensing		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
									A070909	Oversensing		
9475	2019	All other reportable incidents	47265	Dual-chamber implantable pacemaker, rate-responsive	CT141	Electrodes/Leads and associated devices	F1905	Device Revision or Replacement	A26	Insufficient Device Problem Information		
					CT145	Cardiovascular devices	F1905	Device Revision or Replacement	A26	Insufficient Device Problem Information		
					CT1684	Cardiac pacemakers and associated devices	F1905	Device Revision or Replacement	A26	Insufficient Device Problem Information		
					CT1685	Cardiac pacemakers	F1905	Device Revision or Replacement	A26	Insufficient Device Problem Information		
					CT233	Surgical	F1905	Device Revision or Replacement	A26	Insufficient Device Problem Information		
					CT2406	Implantable	F1905	Device Revision or Replacement	A26	Insufficient Device Problem Information		
					CT2407	Active implantable	F1905	Device Revision or Replacement	A26	Insufficient Device Problem Information		
					CT257	Implantable cardiac pacemakers	F1905	Device Revision or Replacement	A26	Insufficient Device Problem Information		
					CT2727	Active implant devices/ancillaries	F1905	Device Revision or Replacement	A26	Insufficient Device Problem Information		
					CT2750	Implantable cardiac electrical stimulators and associated devices	F1905	Device Revision or Replacement	A26	Insufficient Device Problem Information		
					CT2866	Implantable cardiac electrical stimulators	F1905	Device Revision or Replacement	A26	Insufficient Device Problem Information		
					CT2985	Active	F1905	Device Revision or Replacement	A26	Insufficient Device Problem Information		
					CT3001	Implantable metal	F1905	Device Revision or Replacement	A26	Insufficient Device Problem Information		
					CT3002	Implantable synthetic polymer	F1905	Device Revision or Replacement	A26	Insufficient Device Problem Information		
					CT3006	Implantable inorganic material	F1905	Device Revision or Replacement	A26	Insufficient Device Problem Information		
					CT314	Electrically powered	F1905	Device Revision or Replacement	A26	Insufficient Device Problem Information		
					CT321	Long-term surgical invasive	F1905	Device Revision or Replacement	A26	Insufficient Device Problem Information		
					CT449	Generators and associated devices	F1905	Device Revision or Replacement	A26	Insufficient Device Problem Information		
					CT981	Single use	F1905	Device Revision or Replacement	A26	Insufficient Device Problem Information		
					CT983	Surgical invasive	F1905	Device Revision or Replacement	A26	Insufficient Device Problem Information		
9476	2019	All other reportable incidents	47265	Dual-chamber implantable pacemaker, rate-responsive	CT141	Electrodes/Leads and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070101	Failure to Capture		
					CT145	Cardiovascular devices	E2403	No Clinical Signs, Symptoms or Conditions	A070101	Failure to Capture		
					CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070101	Failure to Capture		
					CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A070101	Failure to Capture		
					CT233	Surgical	E2403	No Clinical Signs, Symptoms or Conditions	A070101	Failure to Capture		
					CT2406	Implantable	E2403	No Clinical Signs, Symptoms or Conditions	A070101	Failure to Capture		
					CT2407	Active implantable	E2403	No Clinical Signs, Symptoms or Conditions	A070101	Failure to Capture		
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A070101	Failure to Capture		
					CT2727	Active implant devices/ancillaries	E2403	No Clinical Signs, Symptoms or Conditions	A070101	Failure to Capture		
					CT2750	Implantable cardiac electrical stimulators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070101	Failure to Capture		
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A070101	Failure to Capture		
					CT2985	Active	E2403	No Clinical Signs, Symptoms or Conditions	A070101	Failure to Capture		
					CT3001	Implantable metal	E2403	No Clinical Signs, Symptoms or Conditions	A070101	Failure to Capture		
					CT3002	Implantable synthetic polymer	E2403	No Clinical Signs, Symptoms or Conditions	A070101	Failure to Capture		
					CT3006	Implantable inorganic material	E2403	No Clinical Signs, Symptoms or Conditions	A070101	Failure to Capture		
					CT314	Electrically powered	E2403	No Clinical Signs, Symptoms or Conditions	A070101	Failure to Capture		

					CT321	Long-term surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A070101	Failure to Capture		
					CT449	Generators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070101	Failure to Capture		
					CT981	Single use	E2403	No Clinical Signs, Symptoms or Conditions	A070101	Failure to Capture		
					CT983	Surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A070101	Failure to Capture		
9477	2019	All other reportable incidents	35852	Single-chamber implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A1304	Telemetry Discrepancy		
9478	2019	All other reportable incidents	36241	Pacemaker/defibrillator lead	CT1684	Cardiac pacemakers and associated devices	F1905	Device Revision or Replacement	A072201	High Impedance	17/07/2019	
9479	2019	All other reportable incidents	47270	Cardiac resynchronization therapy implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A070101	Failure to Capture		
9480	2019	Death	17882	Physiologic-monitoring defibrillation system	CT1684	Cardiac pacemakers and associated devices	F21	Unexpected Deterioration	A23	Use of Device Problem		
9481	2019	All other reportable incidents	35852	Single-chamber implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A070909	Oversensing		
9482	2020	All other reportable incidents	37265	Dual-chamber implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	09/07/2020	09/07/2020
							F1901	Additional Surgery	A070504	Premature Discharge of Battery	09/07/2020	09/07/2020
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	09/07/2020	09/07/2020
9483	2020	All other reportable incidents	37265	Dual-chamber implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	03/07/2020	03/07/2020
							F1901	Additional Surgery	A070504	Premature Discharge of Battery	03/07/2020	03/07/2020
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	03/07/2020	03/07/2020
9484	2020	All other reportable incidents	47270	Cardiac resynchronization therapy implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A072201	High Impedance		
							F26	No Health Consequences or Impact	A072201	High Impedance		
9485	2020	All other reportable incidents	47267	Single-chamber implantable pacemaker, rate-responsive	CT141	Electrodes/Leads and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	17/06/2020	17/06/2020
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	17/06/2020	17/06/2020
							F12	Serious Injury/ Illness/ Impairment	A070504	Premature Discharge of Battery	17/06/2020	17/06/2020
							F19	Surgical Intervention	A070504	Premature Discharge of Battery	17/06/2020	17/06/2020
					CT145	Cardiovascular devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	17/06/2020	17/06/2020
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	17/06/2020	17/06/2020
							F12	Serious Injury/ Illness/ Impairment	A070504	Premature Discharge of Battery	17/06/2020	17/06/2020
							F19	Surgical Intervention	A070504	Premature Discharge of Battery	17/06/2020	17/06/2020
					CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	17/06/2020	17/06/2020
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	17/06/2020	17/06/2020
							F12	Serious Injury/ Illness/ Impairment	A070504	Premature Discharge of Battery	17/06/2020	17/06/2020
							F19	Surgical Intervention	A070504	Premature Discharge of Battery	17/06/2020	17/06/2020
					CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	17/06/2020	17/06/2020
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	17/06/2020	17/06/2020
							F12	Serious Injury/ Illness/ Impairment	A070504	Premature Discharge of Battery	17/06/2020	17/06/2020
							F19	Surgical Intervention	A070504	Premature Discharge of Battery	17/06/2020	17/06/2020
					CT233	Surgical	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	17/06/2020	17/06/2020
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	17/06/2020	17/06/2020
							F12	Serious Injury/ Illness/ Impairment	A070504	Premature Discharge of Battery	17/06/2020	17/06/2020
							F19	Surgical Intervention	A070504	Premature Discharge of Battery	17/06/2020	17/06/2020
					CT2406	Implantable	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	17/06/2020	17/06/2020
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	17/06/2020	17/06/2020
							F12	Serious Injury/ Illness/ Impairment	A070504	Premature Discharge of Battery	17/06/2020	17/06/2020
							F19	Surgical Intervention	A070504	Premature Discharge of Battery	17/06/2020	17/06/2020
					CT2407	Active implantable	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	17/06/2020	17/06/2020
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	17/06/2020	17/06/2020
							F12	Serious Injury/ Illness/ Impairment	A070504	Premature Discharge of Battery	17/06/2020	17/06/2020
							F19	Surgical Intervention	A070504	Premature Discharge of Battery	17/06/2020	17/06/2020
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	17/06/2020	17/06/2020
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	17/06/2020	17/06/2020
							F12	Serious Injury/ Illness/ Impairment	A070504	Premature Discharge of Battery	17/06/2020	17/06/2020
							F19	Surgical Intervention	A070504	Premature Discharge of Battery	17/06/2020	17/06/2020
					CT2727	Active implant devices/ancillaries	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	17/06/2020	17/06/2020
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	17/06/2020	17/06/2020
							F12	Serious Injury/ Illness/ Impairment	A070504	Premature Discharge of Battery	17/06/2020	17/06/2020
							F19	Surgical Intervention	A070504	Premature Discharge of Battery	17/06/2020	17/06/2020
					CT2750	Implantable cardiac electrical stimulators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	17/06/2020	17/06/2020
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	17/06/2020	17/06/2020
							F12	Serious Injury/ Illness/ Impairment	A070504	Premature Discharge of Battery	17/06/2020	17/06/2020
							F19	Surgical Intervention	A070504	Premature Discharge of Battery	17/06/2020	17/06/2020
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	17/06/2020	17/06/2020
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	17/06/2020	17/06/2020
							F12	Serious Injury/ Illness/ Impairment	A070504	Premature Discharge of Battery	17/06/2020	17/06/2020
							F19	Surgical Intervention	A070504	Premature Discharge of Battery	17/06/2020	17/06/2020
					CT2985	Active	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	17/06/2020	17/06/2020
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	17/06/2020	17/06/2020
							F12	Serious Injury/ Illness/ Impairment	A070504	Premature Discharge of Battery	17/06/2020	17/06/2020
							F19	Surgical Intervention	A070504	Premature Discharge of Battery	17/06/2020	17/06/2020
					CT3001	Implantable metal	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	17/06/2020	17/06/2020
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	17/06/2020	17/06/2020
							F12	Serious Injury/ Illness/ Impairment	A070504	Premature Discharge of Battery	17/06/2020	17/06/2020
							F19	Surgical Intervention	A070504	Premature Discharge of Battery	17/06/2020	17/06/2020
					CT3002	Implantable synthetic polymer	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	17/06/2020	17/06/2020
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	17/06/2020	17/06/2020
							F12	Serious Injury/ Illness/ Impairment	A070504	Premature Discharge of Battery	17/06/2020	17/06/2020
							F19	Surgical Intervention	A070504	Premature Discharge of Battery	17/06/2020	17/06/2020
					CT3006	Implantable inorganic material	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	17/06/2020	17/06/2020
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	17/06/2020	17/06/2020
							F12	Serious Injury/ Illness/ Impairment	A070504	Premature Discharge of Battery	17/06/2020	17/06/2020
							F19	Surgical Intervention	A070504	Premature Discharge of Battery	17/06/2020	17/06/2020
					CT314	Electrically powered	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	17/06/2020	17/06/2020
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	17/06/2020	17/06/2020
							F12	Serious Injury/ Illness/ Impairment	A070504	Premature Discharge of Battery	17/06/2020	17/06/2020
							F19	Surgical Intervention	A070504	Premature Discharge of Battery	17/06/2020	17/06/2020
					CT321	Long-term surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	17/06/2020	17/06/2020
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	17/06/2020	17/06/2020
							F12	Serious Injury/ Illness/ Impairment	A070504	Premature Discharge of Battery	17/06/2020	17/06/2020
							F19	Surgical Intervention	A070504	Premature Discharge of Battery	17/06/2020	17/06/2020
					CT449	Generators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	17/06/2020	17/06/2020
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	17/06/2020	17/06/2020

							F12	Serious Injury/ Illness/ Impairment	A070504	Premature Discharge of Battery	17/06/2020	17/06/2020
							F19	Surgical Intervention	A070504	Premature Discharge of Battery	17/06/2020	17/06/2020
					CT981	Single use	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	17/06/2020	17/06/2020
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	17/06/2020	17/06/2020
							F12	Serious Injury/ Illness/ Impairment	A070504	Premature Discharge of Battery	17/06/2020	17/06/2020
							F19	Surgical Intervention	A070504	Premature Discharge of Battery	17/06/2020	17/06/2020
					CT983	Surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	17/06/2020	17/06/2020
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	17/06/2020	17/06/2020
							F12	Serious Injury/ Illness/ Impairment	A070504	Premature Discharge of Battery	17/06/2020	17/06/2020
							F19	Surgical Intervention	A070504	Premature Discharge of Battery	17/06/2020	17/06/2020
9486	2020	All other reportable incidents	17882	Physiologic-monitoring defibrillation system	CT1684	Cardiac pacemakers and associated devices	E2401	Insufficient Information	A0719	Unexpected Shutdown		
							E2403	No Clinical Signs, Symptoms or Conditions	A0719	Unexpected Shutdown		
							F26	No Health Consequences or Impact	A0719	Unexpected Shutdown		
9487	2020	All other reportable incidents	35852	Single-chamber implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E2336	Shock	A0709	Device Sensing Problem		
									A070909	Oversensing		
									A071302	Inappropriate/Inadequate Shock/Stimulation		
							F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A0709	Device Sensing Problem		
									A070909	Oversensing		
									A071302	Inappropriate/Inadequate Shock/Stimulation		
							F26	No Health Consequences or Impact	A0709	Device Sensing Problem		
									A070909	Oversensing		
									A071302	Inappropriate/Inadequate Shock/Stimulation		
9488	2019	All other reportable incidents	47265	Dual-chamber implantable pacemaker, rate-responsive	CT141	Electrodes/Leads and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
					CT145	Cardiovascular devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
					CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
					CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
					CT233	Surgical	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
					CT2406	Implantable	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
					CT2407	Active implantable	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
					CT2727	Active implant devices/ancillaries	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
					CT2750	Implantable cardiac electrical stimulators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
					CT2985	Active	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
					CT3001	Implantable metal	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
					CT3002	Implantable synthetic polymer	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
					CT3006	Implantable inorganic material	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
					CT314	Electrically powered	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
					CT321	Long-term surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
					CT449	Generators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
					CT981	Single use	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
					CT983	Surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
9489	2019	All other reportable incidents	37265	Dual-chamber implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	12/08/2019	
9490	2019	All other reportable incidents	36241	Pacemaker/defibrillator lead	CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A05	Mechanical Problem	17/07/2019	17/07/2019
9491	2019	All other reportable incidents	36241	Pacemaker/defibrillator lead	CT1684	Cardiac pacemakers and associated devices	E1906	Unspecified Infection	A24	No Apparent Device Problem	19/07/2019	
							F1905	Device Revision or Replacement	A24	No Apparent Device Problem	19/07/2019	
9492	2019	All other reportable incidents	35852	Single-chamber implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E1906	Unspecified Infection	A24	No Apparent Device Problem	19/07/2019	19/07/2019
							E2006	Erosion	A24	No Apparent Device Problem	19/07/2019	19/07/2019
							F08	Hospitalization or Prolonged Hospitalization	A24	No Apparent Device Problem	19/07/2019	19/07/2019
							F19	Surgical Intervention	A24	No Apparent Device Problem	19/07/2019	19/07/2019
							F1905	Device Revision or Replacement	A24	No Apparent Device Problem	19/07/2019	19/07/2019
9493	2019	All other reportable incidents	37265	Dual-chamber implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	20/08/2019	
9494	2020	All other reportable incidents	17882	Physiologic-monitoring defibrillation system	CT1684	Cardiac pacemakers and associated devices	E2401	Insufficient Information	A071301	Failure to Deliver Shock/Stimulation		
							E2403	No Clinical Signs, Symptoms or Conditions	A071301	Failure to Deliver Shock/Stimulation		
							F24	Insufficient Health Impact Information	A071301	Failure to Deliver Shock/Stimulation		
9495	2020	All other reportable incidents	35223	Endocardial/interventricular septal pacing lead	CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070101	Failure to Capture	02/07/2020	02/07/2020
							F1901	Additional Surgery	A070101	Failure to Capture	02/07/2020	02/07/2020
							F1905	Device Revision or Replacement	A070101	Failure to Capture	02/07/2020	02/07/2020
9496	2020	All other reportable incidents	35852	Single-chamber implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A070909	Oversensing		
							F26	No Health Consequences or Impact	A070909	Oversensing		
9497	2020	All other reportable incidents	37265	Dual-chamber implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A071302	Inappropriate/Inadequate Shock/Stimulation		
							F26	No Health Consequences or Impact	A071302	Inappropriate/Inadequate Shock/Stimulation		
9498	2020	All other reportable incidents	47270	Cardiac resynchronization therapy implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	17/09/2020	17/09/2020
							F19	Surgical Intervention	A070504	Premature Discharge of Battery	17/09/2020	17/09/2020
9499	2020	All other reportable incidents	47265	Dual-chamber implantable pacemaker, rate-responsive	CT141	Electrodes/Leads and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem	07/09/2020	07/09/2020
									A070504	Premature Discharge of Battery	07/09/2020	07/09/2020
							F08	Hospitalization or Prolonged Hospitalization	A0705	Battery Problem	07/09/2020	07/09/2020
									A070504	Premature Discharge of Battery	07/09/2020	07/09/2020
							F12	Serious Injury/ Illness/ Impairment	A0705	Battery Problem	07/09/2020	07/09/2020
									A070504	Premature Discharge of Battery	07/09/2020	07/09/2020
							F19	Surgical Intervention	A0705	Battery Problem	07/09/2020	07/09/2020
									A070504	Premature Discharge of Battery	07/09/2020	07/09/2020
							F1905	Device Revision or Replacement	A0705	Battery Problem	07/09/2020	07/09/2020
									A070504	Premature Discharge of Battery	07/09/2020	07/09/2020
					CT145	Cardiovascular devices	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem	07/09/2020	07/09/2020
									A070504	Premature Discharge of Battery	07/09/2020	07/09/2020
							F08	Hospitalization or Prolonged Hospitalization	A0705	Battery Problem	07/09/2020	07/09/2020
									A070504	Premature Discharge of Battery	07/09/2020	07/09/2020
							F12	Serious Injury/ Illness/ Impairment	A0705	Battery Problem	07/09/2020	07/09/2020
									A070504	Premature Discharge of Battery	07/09/2020	07/09/2020
							F19	Surgical Intervention	A0705	Battery Problem	07/09/2020	07/09/2020
									A070504	Premature Discharge of Battery	07/09/2020	07/09/2020
							F1905	Device Revision or Replacement	A0705	Battery Problem	07/09/2020	07/09/2020
									A070504	Premature Discharge of Battery	07/09/2020	07/09/2020
					CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem	07/09/2020	07/09/2020
									A070504	Premature Discharge of Battery	07/09/2020	07/09/2020
							F08	Hospitalization or Prolonged Hospitalization	A0705	Battery Problem	07/09/2020	07/09/2020

								A070504	Premature Discharge of Battery	07/09/2020	07/09/2020
						F12	Serious Injury/ Illness/ Impairment	A0705	Battery Problem	07/09/2020	07/09/2020
								A070504	Premature Discharge of Battery	07/09/2020	07/09/2020
						F19	Surgical Intervention	A0705	Battery Problem	07/09/2020	07/09/2020
								A070504	Premature Discharge of Battery	07/09/2020	07/09/2020
						F1905	Device Revision or Replacement	A0705	Battery Problem	07/09/2020	07/09/2020
								A070504	Premature Discharge of Battery	07/09/2020	07/09/2020
					CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem	07/09/2020	07/09/2020
									A070504	Premature Discharge of Battery	07/09/2020	07/09/2020
							F08	Hospitalization or Prolonged Hospitalization	A0705	Battery Problem	07/09/2020	07/09/2020
									A070504	Premature Discharge of Battery	07/09/2020	07/09/2020
							F12	Serious Injury/ Illness/ Impairment	A0705	Battery Problem	07/09/2020	07/09/2020
									A070504	Premature Discharge of Battery	07/09/2020	07/09/2020
							F19	Surgical Intervention	A0705	Battery Problem	07/09/2020	07/09/2020
									A070504	Premature Discharge of Battery	07/09/2020	07/09/2020
							F1905	Device Revision or Replacement	A0705	Battery Problem	07/09/2020	07/09/2020
									A070504	Premature Discharge of Battery	07/09/2020	07/09/2020
					CT233	Surgical	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem	07/09/2020	07/09/2020
									A070504	Premature Discharge of Battery	07/09/2020	07/09/2020
							F08	Hospitalization or Prolonged Hospitalization	A0705	Battery Problem	07/09/2020	07/09/2020
									A070504	Premature Discharge of Battery	07/09/2020	07/09/2020
							F12	Serious Injury/ Illness/ Impairment	A0705	Battery Problem	07/09/2020	07/09/2020
									A070504	Premature Discharge of Battery	07/09/2020	07/09/2020
							F19	Surgical Intervention	A0705	Battery Problem	07/09/2020	07/09/2020
									A070504	Premature Discharge of Battery	07/09/2020	07/09/2020
							F1905	Device Revision or Replacement	A0705	Battery Problem	07/09/2020	07/09/2020
									A070504	Premature Discharge of Battery	07/09/2020	07/09/2020
					CT2406	Implantable	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem	07/09/2020	07/09/2020
									A070504	Premature Discharge of Battery	07/09/2020	07/09/2020
							F08	Hospitalization or Prolonged Hospitalization	A0705	Battery Problem	07/09/2020	07/09/2020
									A070504	Premature Discharge of Battery	07/09/2020	07/09/2020
							F12	Serious Injury/ Illness/ Impairment	A0705	Battery Problem	07/09/2020	07/09/2020
									A070504	Premature Discharge of Battery	07/09/2020	07/09/2020
							F19	Surgical Intervention	A0705	Battery Problem	07/09/2020	07/09/2020
									A070504	Premature Discharge of Battery	07/09/2020	07/09/2020
							F1905	Device Revision or Replacement	A0705	Battery Problem	07/09/2020	07/09/2020
									A070504	Premature Discharge of Battery	07/09/2020	07/09/2020
					CT2407	Active implantable	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem	07/09/2020	07/09/2020
									A070504	Premature Discharge of Battery	07/09/2020	07/09/2020
							F08	Hospitalization or Prolonged Hospitalization	A0705	Battery Problem	07/09/2020	07/09/2020
									A070504	Premature Discharge of Battery	07/09/2020	07/09/2020
							F12	Serious Injury/ Illness/ Impairment	A0705	Battery Problem	07/09/2020	07/09/2020
									A070504	Premature Discharge of Battery	07/09/2020	07/09/2020
							F19	Surgical Intervention	A0705	Battery Problem	07/09/2020	07/09/2020
									A070504	Premature Discharge of Battery	07/09/2020	07/09/2020
							F1905	Device Revision or Replacement	A0705	Battery Problem	07/09/2020	07/09/2020
									A070504	Premature Discharge of Battery	07/09/2020	07/09/2020
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem	07/09/2020	07/09/2020
									A070504	Premature Discharge of Battery	07/09/2020	07/09/2020
							F08	Hospitalization or Prolonged Hospitalization	A0705	Battery Problem	07/09/2020	07/09/2020
									A070504	Premature Discharge of Battery	07/09/2020	07/09/2020
							F12	Serious Injury/ Illness/ Impairment	A0705	Battery Problem	07/09/2020	07/09/2020
									A070504	Premature Discharge of Battery	07/09/2020	07/09/2020
							F19	Surgical Intervention	A0705	Battery Problem	07/09/2020	07/09/2020
									A070504	Premature Discharge of Battery	07/09/2020	07/09/2020
							F1905	Device Revision or Replacement	A0705	Battery Problem	07/09/2020	07/09/2020
									A070504	Premature Discharge of Battery	07/09/2020	07/09/2020
					CT2727	Active implant devices/ancillaries	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem	07/09/2020	07/09/2020
									A070504	Premature Discharge of Battery	07/09/2020	07/09/2020
							F08	Hospitalization or Prolonged Hospitalization	A0705	Battery Problem	07/09/2020	07/09/2020
									A070504	Premature Discharge of Battery	07/09/2020	07/09/2020
							F12	Serious Injury/ Illness/ Impairment	A0705	Battery Problem	07/09/2020	07/09/2020
									A070504	Premature Discharge of Battery	07/09/2020	07/09/2020
							F19	Surgical Intervention	A0705	Battery Problem	07/09/2020	07/09/2020
									A070504	Premature Discharge of Battery	07/09/2020	07/09/2020
							F1905	Device Revision or Replacement	A0705	Battery Problem	07/09/2020	07/09/2020
									A070504	Premature Discharge of Battery	07/09/2020	07/09/2020
					CT2750	Implantable cardiac electrical stimulators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem	07/09/2020	07/09/2020
									A070504	Premature Discharge of Battery	07/09/2020	07/09/2020
							F08	Hospitalization or Prolonged Hospitalization	A0705	Battery Problem	07/09/2020	07/09/2020
									A070504	Premature Discharge of Battery	07/09/2020	07/09/2020
							F12	Serious Injury/ Illness/ Impairment	A0705	Battery Problem	07/09/2020	07/09/2020
									A070504	Premature Discharge of Battery	07/09/2020	07/09/2020
							F19	Surgical Intervention	A0705	Battery Problem	07/09/2020	07/09/2020
									A070504	Premature Discharge of Battery	07/09/2020	07/09/2020
							F1905	Device Revision or Replacement	A0705	Battery Problem	07/09/2020	07/09/2020
									A070504	Premature Discharge of Battery	07/09/2020	07/09/2020
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem	07/09/2020	07/09/2020
									A070504	Premature Discharge of Battery	07/09/2020	07/09/2020
							F08	Hospitalization or Prolonged Hospitalization	A0705	Battery Problem	07/09/2020	07/09/2020
									A070504	Premature Discharge of Battery	07/09/2020	07/09/2020
							F12	Serious Injury/ Illness/ Impairment	A0705	Battery Problem	07/09/2020	07/09/2020
									A070504	Premature Discharge of Battery	07/09/2020	07/09/2020
							F19	Surgical Intervention	A0705	Battery Problem	07/09/2020	07/09/2020
									A070504	Premature Discharge of Battery	07/09/2020	07/09/2020
							F1905	Device Revision or Replacement	A0705	Battery Problem	07/09/2020	07/09/2020
									A070504	Premature Discharge of Battery	07/09/2020	07/09/2020
					CT2985	Active	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem	07/09/2020	07/09/2020

								A070504	Premature Discharge of Battery	07/09/2020	07/09/2020
						F08	Hospitalization or Prolonged Hospitalization	A0705	Battery Problem	07/09/2020	07/09/2020
								A070504	Premature Discharge of Battery	07/09/2020	07/09/2020
						F12	Serious Injury/ Illness/ Impairment	A0705	Battery Problem	07/09/2020	07/09/2020
								A070504	Premature Discharge of Battery	07/09/2020	07/09/2020
						F19	Surgical Intervention	A0705	Battery Problem	07/09/2020	07/09/2020
								A070504	Premature Discharge of Battery	07/09/2020	07/09/2020
						F1905	Device Revision or Replacement	A0705	Battery Problem	07/09/2020	07/09/2020
								A070504	Premature Discharge of Battery	07/09/2020	07/09/2020
					CT3001	Implantable metal	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem	07/09/2020	07/09/2020
									A070504	Premature Discharge of Battery	07/09/2020	07/09/2020
							F08	Hospitalization or Prolonged Hospitalization	A0705	Battery Problem	07/09/2020	07/09/2020
									A070504	Premature Discharge of Battery	07/09/2020	07/09/2020
							F12	Serious Injury/ Illness/ Impairment	A0705	Battery Problem	07/09/2020	07/09/2020
									A070504	Premature Discharge of Battery	07/09/2020	07/09/2020
							F19	Surgical Intervention	A0705	Battery Problem	07/09/2020	07/09/2020
									A070504	Premature Discharge of Battery	07/09/2020	07/09/2020
							F1905	Device Revision or Replacement	A0705	Battery Problem	07/09/2020	07/09/2020
									A070504	Premature Discharge of Battery	07/09/2020	07/09/2020
					CT3002	Implantable synthetic polymer	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem	07/09/2020	07/09/2020
									A070504	Premature Discharge of Battery	07/09/2020	07/09/2020
							F08	Hospitalization or Prolonged Hospitalization	A0705	Battery Problem	07/09/2020	07/09/2020
									A070504	Premature Discharge of Battery	07/09/2020	07/09/2020
							F12	Serious Injury/ Illness/ Impairment	A0705	Battery Problem	07/09/2020	07/09/2020
									A070504	Premature Discharge of Battery	07/09/2020	07/09/2020
							F19	Surgical Intervention	A0705	Battery Problem	07/09/2020	07/09/2020
									A070504	Premature Discharge of Battery	07/09/2020	07/09/2020
							F1905	Device Revision or Replacement	A0705	Battery Problem	07/09/2020	07/09/2020
									A070504	Premature Discharge of Battery	07/09/2020	07/09/2020
					CT3006	Implantable inorganic material	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem	07/09/2020	07/09/2020
									A070504	Premature Discharge of Battery	07/09/2020	07/09/2020
							F08	Hospitalization or Prolonged Hospitalization	A0705	Battery Problem	07/09/2020	07/09/2020
									A070504	Premature Discharge of Battery	07/09/2020	07/09/2020
							F12	Serious Injury/ Illness/ Impairment	A0705	Battery Problem	07/09/2020	07/09/2020
									A070504	Premature Discharge of Battery	07/09/2020	07/09/2020
							F19	Surgical Intervention	A0705	Battery Problem	07/09/2020	07/09/2020
									A070504	Premature Discharge of Battery	07/09/2020	07/09/2020
							F1905	Device Revision or Replacement	A0705	Battery Problem	07/09/2020	07/09/2020
									A070504	Premature Discharge of Battery	07/09/2020	07/09/2020
					CT314	Electrically powered	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem	07/09/2020	07/09/2020
									A070504	Premature Discharge of Battery	07/09/2020	07/09/2020
							F08	Hospitalization or Prolonged Hospitalization	A0705	Battery Problem	07/09/2020	07/09/2020
									A070504	Premature Discharge of Battery	07/09/2020	07/09/2020
							F12	Serious Injury/ Illness/ Impairment	A0705	Battery Problem	07/09/2020	07/09/2020
									A070504	Premature Discharge of Battery	07/09/2020	07/09/2020
							F19	Surgical Intervention	A0705	Battery Problem	07/09/2020	07/09/2020
									A070504	Premature Discharge of Battery	07/09/2020	07/09/2020
							F1905	Device Revision or Replacement	A0705	Battery Problem	07/09/2020	07/09/2020
									A070504	Premature Discharge of Battery	07/09/2020	07/09/2020
					CT321	Long-term surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem	07/09/2020	07/09/2020
									A070504	Premature Discharge of Battery	07/09/2020	07/09/2020
							F08	Hospitalization or Prolonged Hospitalization	A0705	Battery Problem	07/09/2020	07/09/2020
									A070504	Premature Discharge of Battery	07/09/2020	07/09/2020
							F12	Serious Injury/ Illness/ Impairment	A0705	Battery Problem	07/09/2020	07/09/2020
									A070504	Premature Discharge of Battery	07/09/2020	07/09/2020
							F19	Surgical Intervention	A0705	Battery Problem	07/09/2020	07/09/2020
									A070504	Premature Discharge of Battery	07/09/2020	07/09/2020
							F1905	Device Revision or Replacement	A0705	Battery Problem	07/09/2020	07/09/2020
									A070504	Premature Discharge of Battery	07/09/2020	07/09/2020
					CT449	Generators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem	07/09/2020	07/09/2020
									A070504	Premature Discharge of Battery	07/09/2020	07/09/2020
							F08	Hospitalization or Prolonged Hospitalization	A0705	Battery Problem	07/09/2020	07/09/2020
									A070504	Premature Discharge of Battery	07/09/2020	07/09/2020
							F12	Serious Injury/ Illness/ Impairment	A0705	Battery Problem	07/09/2020	07/09/2020
									A070504	Premature Discharge of Battery	07/09/2020	07/09/2020
							F19	Surgical Intervention	A0705	Battery Problem	07/09/2020	07/09/2020
									A070504	Premature Discharge of Battery	07/09/2020	07/09/2020
							F1905	Device Revision or Replacement	A0705	Battery Problem	07/09/2020	07/09/2020
									A070504	Premature Discharge of Battery	07/09/2020	07/09/2020
					CT981	Single use	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem	07/09/2020	07/09/2020
									A070504	Premature Discharge of Battery	07/09/2020	07/09/2020
							F08	Hospitalization or Prolonged Hospitalization	A0705	Battery Problem	07/09/2020	07/09/2020
									A070504	Premature Discharge of Battery	07/09/2020	07/09/2020
							F12	Serious Injury/ Illness/ Impairment	A0705	Battery Problem	07/09/2020	07/09/2020
									A070504	Premature Discharge of Battery	07/09/2020	07/09/2020
							F19	Surgical Intervention	A0705	Battery Problem	07/09/2020	07/09/2020
									A070504	Premature Discharge of Battery	07/09/2020	07/09/2020
							F1905	Device Revision or Replacement	A0705	Battery Problem	07/09/2020	07/09/2020
									A070504	Premature Discharge of Battery	07/09/2020	07/09/2020
					CT983	Surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem	07/09/2020	07/09/2020
									A070504	Premature Discharge of Battery	07/09/2020	07/09/2020
							F08	Hospitalization or Prolonged Hospitalization	A0705	Battery Problem	07/09/2020	07/09/2020
									A070504	Premature Discharge of Battery	07/09/2020	07/09/2020
							F12	Serious Injury/ Illness/ Impairment	A0705	Battery Problem	07/09/2020	07/09/2020
									A070504	Premature Discharge of Battery	07/09/2020	07/09/2020
							F19	Surgical Intervention	A0705	Battery Problem	07/09/2020	07/09/2020
									A070504	Premature Discharge of Battery	07/09/2020	07/09/2020
							F1905	Device Revision or Replacement	A0705	Battery Problem	07/09/2020	07/09/2020

									A070504	Premature Discharge of Battery	07/09/2020	07/09/2020
9500	2019	All other reportable incidents	60789	Intracardiac pacemaker	CT130	Catheters and associated devices	F14	Prolonged Episode of Care	A071204	No Pacing	04/06/2019	
					CT145	Cardiovascular devices	F14	Prolonged Episode of Care	A071204	No Pacing	04/06/2019	
					CT1684	Cardiac pacemakers and associated devices	F14	Prolonged Episode of Care	A071204	No Pacing	04/06/2019	
					CT1685	Cardiac pacemakers	F14	Prolonged Episode of Care	A071204	No Pacing	04/06/2019	
					CT233	Surgical	F14	Prolonged Episode of Care	A071204	No Pacing	04/06/2019	
					CT2406	Implantable	F14	Prolonged Episode of Care	A071204	No Pacing	04/06/2019	
					CT2407	Active implantable	F14	Prolonged Episode of Care	A071204	No Pacing	04/06/2019	
					CT257	Implantable cardiac pacemakers	F14	Prolonged Episode of Care	A071204	No Pacing	04/06/2019	
					CT2727	Active implant devices/ancillaries	F14	Prolonged Episode of Care	A071204	No Pacing	04/06/2019	
					CT2750	Implantable cardiac electrical stimulators and associated devices	F14	Prolonged Episode of Care	A071204	No Pacing	04/06/2019	
					CT2818	Catheterization/Cannulation/Fluid path devices	F14	Prolonged Episode of Care	A071204	No Pacing	04/06/2019	
					CT2866	Implantable cardiac electrical stimulators	F14	Prolonged Episode of Care	A071204	No Pacing	04/06/2019	
					CT2985	Active	F14	Prolonged Episode of Care	A071204	No Pacing	04/06/2019	
					CT3001	Implantable metal	F14	Prolonged Episode of Care	A071204	No Pacing	04/06/2019	
					CT3002	Implantable synthetic polymer	F14	Prolonged Episode of Care	A071204	No Pacing	04/06/2019	
					CT3006	Implantable inorganic material	F14	Prolonged Episode of Care	A071204	No Pacing	04/06/2019	
					CT314	Electrically powered	F14	Prolonged Episode of Care	A071204	No Pacing	04/06/2019	
					CT321	Long-term surgical invasive	F14	Prolonged Episode of Care	A071204	No Pacing	04/06/2019	
					CT981	Single use	F14	Prolonged Episode of Care	A071204	No Pacing	04/06/2019	
					CT983	Surgical invasive	F14	Prolonged Episode of Care	A071204	No Pacing	04/06/2019	
9501	2019	All other reportable incidents	35223	Endocardial/interventricular septal pacing lead	CT1684	Cardiac pacemakers and associated devices	F1905	Device Revision or Replacement	A0709	Device Sensing Problem	31/07/2019	
9502	2019	All other reportable incidents	35852	Single-chamber implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A071302	Inappropriate/Inadequate Shock/Stimulation	26/09/2019	26/09/2019
9503	2019	All other reportable incidents	36241	Pacemaker/defibrillator lead	CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A0510	Retraction Problem	22/07/2019	
9504	2019	All other reportable incidents	36241	Pacemaker/defibrillator lead	CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A0510	Retraction Problem	22/07/2019	
9505	2020	All other reportable incidents	35223	Endocardial/interventricular septal pacing lead	CT1684	Cardiac pacemakers and associated devices	E2401	Insufficient Information	A0103	Loosening of Implant not Related to Bone-Ingrowth		
							F1904	Device Repositioning	A0103	Loosening of Implant not Related to Bone-Ingrowth		
							F1905	Device Revision or Replacement	A0103	Loosening of Implant not Related to Bone-Ingrowth		
9506	2019	All other reportable incidents	47265	Dual-chamber implantable pacemaker, rate-responsive	CT141	Electrodes/Leads and associated devices	F21	Unexpected Deterioration	A24	No Apparent Device Problem		
					CT145	Cardiovascular devices	F21	Unexpected Deterioration	A24	No Apparent Device Problem		
					CT1684	Cardiac pacemakers and associated devices	F21	Unexpected Deterioration	A24	No Apparent Device Problem		
					CT1685	Cardiac pacemakers	F21	Unexpected Deterioration	A24	No Apparent Device Problem		
					CT233	Surgical	F21	Unexpected Deterioration	A24	No Apparent Device Problem		
					CT2406	Implantable	F21	Unexpected Deterioration	A24	No Apparent Device Problem		
					CT2407	Active implantable	F21	Unexpected Deterioration	A24	No Apparent Device Problem		
					CT257	Implantable cardiac pacemakers	F21	Unexpected Deterioration	A24	No Apparent Device Problem		
					CT2727	Active implant devices/ancillaries	F21	Unexpected Deterioration	A24	No Apparent Device Problem		
					CT2750	Implantable cardiac electrical stimulators and associated devices	F21	Unexpected Deterioration	A24	No Apparent Device Problem		
					CT2866	Implantable cardiac electrical stimulators	F21	Unexpected Deterioration	A24	No Apparent Device Problem		
					CT2985	Active	F21	Unexpected Deterioration	A24	No Apparent Device Problem		
					CT3001	Implantable metal	F21	Unexpected Deterioration	A24	No Apparent Device Problem		
					CT3002	Implantable synthetic polymer	F21	Unexpected Deterioration	A24	No Apparent Device Problem		
					CT3006	Implantable inorganic material	F21	Unexpected Deterioration	A24	No Apparent Device Problem		
					CT314	Electrically powered	F21	Unexpected Deterioration	A24	No Apparent Device Problem		
					CT321	Long-term surgical invasive	F21	Unexpected Deterioration	A24	No Apparent Device Problem		
					CT449	Generators and associated devices	F21	Unexpected Deterioration	A24	No Apparent Device Problem		
					CT981	Single use	F21	Unexpected Deterioration	A24	No Apparent Device Problem		
					CT983	Surgical invasive	F21	Unexpected Deterioration	A24	No Apparent Device Problem		
9507	2019	All other reportable incidents	47265	Dual-chamber implantable pacemaker, rate-responsive	CT141	Electrodes/Leads and associated devices	F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	17/05/2019	
					CT145	Cardiovascular devices	F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	17/05/2019	
					CT1684	Cardiac pacemakers and associated devices	F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	17/05/2019	
					CT1685	Cardiac pacemakers	F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	17/05/2019	
					CT233	Surgical	F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	17/05/2019	
					CT2406	Implantable	F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	17/05/2019	
					CT2407	Active implantable	F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	17/05/2019	
					CT257	Implantable cardiac pacemakers	F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	17/05/2019	
					CT2727	Active implant devices/ancillaries	F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	17/05/2019	
					CT2750	Implantable cardiac electrical stimulators and associated devices	F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	17/05/2019	
					CT2866	Implantable cardiac electrical stimulators	F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	17/05/2019	
					CT2985	Active	F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	17/05/2019	
					CT3001	Implantable metal	F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	17/05/2019	
					CT3002	Implantable synthetic polymer	F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	17/05/2019	
					CT3006	Implantable inorganic material	F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	17/05/2019	
					CT314	Electrically powered	F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	17/05/2019	
					CT321	Long-term surgical invasive	F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	17/05/2019	
					CT449	Generators and associated devices	F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	17/05/2019	
					CT981	Single use	F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	17/05/2019	
					CT983	Surgical invasive	F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	17/05/2019	
9508	2020	All other reportable incidents	47263	Cardiac resynchronization therapy implantable pacemaker	CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A071102	Failure to Interrogate	26/06/2020	26/06/2020
									A1305	Wireless Communication Problem	26/06/2020	26/06/2020
							F1901	Additional Surgery	A071102	Failure to Interrogate	26/06/2020	26/06/2020
									A1305	Wireless Communication Problem	26/06/2020	26/06/2020
							F1905	Device Revision or Replacement	A071102	Failure to Interrogate	26/06/2020	26/06/2020
									A1305	Wireless Communication Problem	26/06/2020	26/06/2020
					CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A071102	Failure to Interrogate	26/06/2020	26/06/2020
									A1305	Wireless Communication Problem	26/06/2020	26/06/2020
							F1901	Additional Surgery	A071102	Failure to Interrogate	26/06/2020	26/06/2020
									A1305	Wireless Communication Problem	26/06/2020	26/06/2020
							F1905	Device Revision or Replacement	A071102	Failure to Interrogate	26/06/2020	26/06/2020
									A1305	Wireless Communication Problem	26/06/2020	26/06/2020
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A071102	Failure to Interrogate	26/06/2020	26/06/2020
									A1305	Wireless Communication Problem	26/06/2020	26/06/2020

							F1901	Additional Surgery	A071102	Failure to Interrogate	26/06/2020	26/06/2020
									A1305	Wireless Communication Problem	26/06/2020	26/06/2020
							F1905	Device Revision or Replacement	A071102	Failure to Interrogate	26/06/2020	26/06/2020
									A1305	Wireless Communication Problem	26/06/2020	26/06/2020
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A071102	Failure to Interrogate	26/06/2020	26/06/2020
									A1305	Wireless Communication Problem	26/06/2020	26/06/2020
							F1901	Additional Surgery	A071102	Failure to Interrogate	26/06/2020	26/06/2020
									A1305	Wireless Communication Problem	26/06/2020	26/06/2020
							F1905	Device Revision or Replacement	A071102	Failure to Interrogate	26/06/2020	26/06/2020
									A1305	Wireless Communication Problem	26/06/2020	26/06/2020
9509	2020	All other reportable incidents	47270	Cardiac resynchronization therapy implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	08/07/2020	08/07/2020
							F1901	Additional Surgery	A070504	Premature Discharge of Battery	08/07/2020	08/07/2020
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	08/07/2020	08/07/2020
9510	2019	All other reportable incidents	47270	Cardiac resynchronization therapy implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	F1905	Device Revision or Replacement	A071204	No Pacing	09/08/2019	
9511	2019	All other reportable incidents	37265	Dual-chamber implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A0203	Defective Device	29/07/2019	29/07/2019
									A090801	Signal Artifact/Noise	29/07/2019	29/07/2019
							F08	Hospitalization or Prolonged Hospitalization	A0203	Defective Device	29/07/2019	29/07/2019
									A090801	Signal Artifact/Noise	29/07/2019	29/07/2019
							F1905	Device Revision or Replacement	A0203	Defective Device	29/07/2019	29/07/2019
									A090801	Signal Artifact/Noise	29/07/2019	29/07/2019
9512	2019	All other reportable incidents	36241	Pacemaker/defibrillator lead	CT1684	Cardiac pacemakers and associated devices	F1905	Device Revision or Replacement	A04	Material Integrity Problem	25/07/2019	25/07/2019
9513	2019	All other reportable incidents	36241	Pacemaker/defibrillator lead	CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A0405	Degraded	25/07/2019	
9514	2020	All other reportable incidents	47265	Dual-chamber implantable pacemaker, rate-responsive	CT141	Electrodes/Leads and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	17/08/2020	17/08/2020
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	17/08/2020	17/08/2020
					CT145	Cardiovascular devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	17/08/2020	17/08/2020
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	17/08/2020	17/08/2020
					CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	17/08/2020	17/08/2020
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	17/08/2020	17/08/2020
					CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	17/08/2020	17/08/2020
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	17/08/2020	17/08/2020
					CT233	Surgical	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	17/08/2020	17/08/2020
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	17/08/2020	17/08/2020
					CT2406	Implantable	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	17/08/2020	17/08/2020
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	17/08/2020	17/08/2020
					CT2407	Active implantable	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	17/08/2020	17/08/2020
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	17/08/2020	17/08/2020
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	17/08/2020	17/08/2020
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	17/08/2020	17/08/2020
					CT2727	Active implant devices/ancillaries	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	17/08/2020	17/08/2020
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	17/08/2020	17/08/2020
					CT2750	Implantable cardiac electrical stimulators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	17/08/2020	17/08/2020
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	17/08/2020	17/08/2020
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	17/08/2020	17/08/2020
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	17/08/2020	17/08/2020
					CT2985	Active	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	17/08/2020	17/08/2020
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	17/08/2020	17/08/2020
					CT3001	Implantable metal	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	17/08/2020	17/08/2020
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	17/08/2020	17/08/2020
					CT3002	Implantable synthetic polymer	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	17/08/2020	17/08/2020
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	17/08/2020	17/08/2020
					CT3006	Implantable inorganic material	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	17/08/2020	17/08/2020
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	17/08/2020	17/08/2020
					CT314	Electrically powered	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	17/08/2020	17/08/2020
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	17/08/2020	17/08/2020
					CT321	Long-term surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	17/08/2020	17/08/2020
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	17/08/2020	17/08/2020
					CT449	Generators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	17/08/2020	17/08/2020
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	17/08/2020	17/08/2020
					CT981	Single use	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	17/08/2020	17/08/2020
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	17/08/2020	17/08/2020
					CT983	Surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	17/08/2020	17/08/2020
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	17/08/2020	17/08/2020
9515	2020	All other reportable incidents	47265	Dual-chamber implantable pacemaker, rate-responsive	CT141	Electrodes/Leads and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	30/09/2020	30/09/2020
									A070909	Oversensing	30/09/2020	30/09/2020
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	30/09/2020	30/09/2020
									A070909	Oversensing	30/09/2020	30/09/2020
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	30/09/2020	30/09/2020
									A070909	Oversensing	30/09/2020	30/09/2020
					CT145	Cardiovascular devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	30/09/2020	30/09/2020
									A070909	Oversensing	30/09/2020	30/09/2020
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	30/09/2020	30/09/2020
									A070909	Oversensing	30/09/2020	30/09/2020
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	30/09/2020	30/09/2020
									A070909	Oversensing	30/09/2020	30/09/2020
					CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	30/09/2020	30/09/2020
									A070909	Oversensing	30/09/2020	30/09/2020
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	30/09/2020	30/09/2020
									A070909	Oversensing	30/09/2020	30/09/2020
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	30/09/2020	30/09/2020
									A070909	Oversensing	30/09/2020	30/09/2020
					CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	30/09/2020	30/09/2020
									A070909	Oversensing	30/09/2020	30/09/2020
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	30/09/2020	30/09/2020
									A070909	Oversensing	30/09/2020	30/09/2020
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	30/09/2020	30/09/2020
									A070909	Oversensing	30/09/2020	30/09/2020
					CT233	Surgical	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	30/09/2020	30/09/2020

									A070909	Oversensing	30/09/2020	30/09/2020
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	30/09/2020	30/09/2020
									A070909	Oversensing	30/09/2020	30/09/2020
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	30/09/2020	30/09/2020
									A070909	Oversensing	30/09/2020	30/09/2020
					CT2406	Implantable	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	30/09/2020	30/09/2020
									A070909	Oversensing	30/09/2020	30/09/2020
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	30/09/2020	30/09/2020
									A070909	Oversensing	30/09/2020	30/09/2020
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	30/09/2020	30/09/2020
									A070909	Oversensing	30/09/2020	30/09/2020
					CT2407	Active implantable	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	30/09/2020	30/09/2020
									A070909	Oversensing	30/09/2020	30/09/2020
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	30/09/2020	30/09/2020
									A070909	Oversensing	30/09/2020	30/09/2020
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	30/09/2020	30/09/2020
									A070909	Oversensing	30/09/2020	30/09/2020
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	30/09/2020	30/09/2020
									A070909	Oversensing	30/09/2020	30/09/2020
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	30/09/2020	30/09/2020
									A070909	Oversensing	30/09/2020	30/09/2020
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	30/09/2020	30/09/2020
									A070909	Oversensing	30/09/2020	30/09/2020
					CT2727	Active implant devices/ancillaries	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	30/09/2020	30/09/2020
									A070909	Oversensing	30/09/2020	30/09/2020
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	30/09/2020	30/09/2020
									A070909	Oversensing	30/09/2020	30/09/2020
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	30/09/2020	30/09/2020
									A070909	Oversensing	30/09/2020	30/09/2020
					CT2750	Implantable cardiac electrical stimulators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	30/09/2020	30/09/2020
									A070909	Oversensing	30/09/2020	30/09/2020
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	30/09/2020	30/09/2020
									A070909	Oversensing	30/09/2020	30/09/2020
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	30/09/2020	30/09/2020
									A070909	Oversensing	30/09/2020	30/09/2020
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	30/09/2020	30/09/2020
									A070909	Oversensing	30/09/2020	30/09/2020
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	30/09/2020	30/09/2020
									A070909	Oversensing	30/09/2020	30/09/2020
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	30/09/2020	30/09/2020
									A070909	Oversensing	30/09/2020	30/09/2020
					CT2985	Active	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	30/09/2020	30/09/2020
									A070909	Oversensing	30/09/2020	30/09/2020
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	30/09/2020	30/09/2020
									A070909	Oversensing	30/09/2020	30/09/2020
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	30/09/2020	30/09/2020
									A070909	Oversensing	30/09/2020	30/09/2020
					CT3001	Implantable metal	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	30/09/2020	30/09/2020
									A070909	Oversensing	30/09/2020	30/09/2020
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	30/09/2020	30/09/2020
									A070909	Oversensing	30/09/2020	30/09/2020
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	30/09/2020	30/09/2020
									A070909	Oversensing	30/09/2020	30/09/2020
					CT3002	Implantable synthetic polymer	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	30/09/2020	30/09/2020
									A070909	Oversensing	30/09/2020	30/09/2020
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	30/09/2020	30/09/2020
									A070909	Oversensing	30/09/2020	30/09/2020
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	30/09/2020	30/09/2020
									A070909	Oversensing	30/09/2020	30/09/2020
					CT3006	Implantable inorganic material	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	30/09/2020	30/09/2020
									A070909	Oversensing	30/09/2020	30/09/2020
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	30/09/2020	30/09/2020
									A070909	Oversensing	30/09/2020	30/09/2020
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	30/09/2020	30/09/2020
									A070909	Oversensing	30/09/2020	30/09/2020
					CT314	Electrically powered	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	30/09/2020	30/09/2020
									A070909	Oversensing	30/09/2020	30/09/2020
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	30/09/2020	30/09/2020
									A070909	Oversensing	30/09/2020	30/09/2020
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	30/09/2020	30/09/2020
									A070909	Oversensing	30/09/2020	30/09/2020
					CT321	Long-term surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	30/09/2020	30/09/2020
									A070909	Oversensing	30/09/2020	30/09/2020
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	30/09/2020	30/09/2020
									A070909	Oversensing	30/09/2020	30/09/2020
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	30/09/2020	30/09/2020
									A070909	Oversensing	30/09/2020	30/09/2020
					CT449	Generators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	30/09/2020	30/09/2020
									A070909	Oversensing	30/09/2020	30/09/2020
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	30/09/2020	30/09/2020
									A070909	Oversensing	30/09/2020	30/09/2020
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	30/09/2020	30/09/2020
									A070909	Oversensing	30/09/2020	30/09/2020
					CT981	Single use	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	30/09/2020	30/09/2020
									A070909	Oversensing	30/09/2020	30/09/2020
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	30/09/2020	30/09/2020
									A070909	Oversensing	30/09/2020	30/09/2020
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	30/09/2020	30/09/2020

									A070909	Oversensing	30/09/2020	30/09/2020
					CT983	Surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	30/09/2020	30/09/2020
									A070909	Oversensing	30/09/2020	30/09/2020
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	30/09/2020	30/09/2020
									A070909	Oversensing	30/09/2020	30/09/2020
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	30/09/2020	30/09/2020
									A070909	Oversensing	30/09/2020	30/09/2020
9516	2020	All other reportable incidents	35852	Single-chamber implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A070909	Oversensing		
							F26	No Health Consequences or Impact	A070909	Oversensing		
9517	2020	All other reportable incidents	36241	Pacemaker/defibrillator lead	CT1684	Cardiac pacemakers and associated devices	E2104	Electric Shock	A070909	Oversensing		
									A071302	Inappropriate/Inadequate Shock/Stimulation		
							F19	Surgical Intervention	A070909	Oversensing		
									A071302	Inappropriate/Inadequate Shock/Stimulation		
							F1905	Device Revision or Replacement	A070909	Oversensing		
									A071302	Inappropriate/Inadequate Shock/Stimulation		
9518	2019	All other reportable incidents	36241	Pacemaker/defibrillator lead	CT1684	Cardiac pacemakers and associated devices	F1905	Device Revision or Replacement	A	Annex A	26/07/2019	
9519	2020	All other reportable incidents	17882	Physiologic-monitoring defibrillation system	CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A0719	Unexpected Shutdown		
							F27	Problem Identified During Non-Clinical Procedure	A0719	Unexpected Shutdown		
9520	2019	All other reportable incidents	36241	Pacemaker/defibrillator lead	CT1684	Cardiac pacemakers and associated devices	F1905	Device Revision or Replacement	A010402	Migration	05/05/2019	
9521	2019	All other reportable incidents	35852	Single-chamber implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	20/08/2019	
9522	2019	All other reportable incidents	47265	Dual-chamber implantable pacemaker, rate-responsive	CT141	Electrodes/Leads and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A072201	High Impedance		
					CT145	Cardiovascular devices	E2403	No Clinical Signs, Symptoms or Conditions	A072201	High Impedance		
					CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A072201	High Impedance		
					CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A072201	High Impedance		
					CT233	Surgical	E2403	No Clinical Signs, Symptoms or Conditions	A072201	High Impedance		
					CT2406	Implantable	E2403	No Clinical Signs, Symptoms or Conditions	A072201	High Impedance		
					CT2407	Active implantable	E2403	No Clinical Signs, Symptoms or Conditions	A072201	High Impedance		
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A072201	High Impedance		
					CT2727	Active implant devices/ancillaries	E2403	No Clinical Signs, Symptoms or Conditions	A072201	High Impedance		
					CT2750	Implantable cardiac electrical stimulators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A072201	High Impedance		
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A072201	High Impedance		
					CT2985	Active	E2403	No Clinical Signs, Symptoms or Conditions	A072201	High Impedance		
					CT3001	Implantable metal	E2403	No Clinical Signs, Symptoms or Conditions	A072201	High Impedance		
					CT3002	Implantable synthetic polymer	E2403	No Clinical Signs, Symptoms or Conditions	A072201	High Impedance		
					CT3006	Implantable inorganic material	E2403	No Clinical Signs, Symptoms or Conditions	A072201	High Impedance		
					CT314	Electrically powered	E2403	No Clinical Signs, Symptoms or Conditions	A072201	High Impedance		
					CT321	Long-term surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A072201	High Impedance		
					CT449	Generators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A072201	High Impedance		
					CT981	Single use	E2403	No Clinical Signs, Symptoms or Conditions	A072201	High Impedance		
					CT983	Surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A072201	High Impedance		
9523	2019	All other reportable incidents	36241	Pacemaker/defibrillator lead	CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A0404	Crack		
9524	2019	All other reportable incidents	36241	Pacemaker/defibrillator lead	CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070909	Oversensing		
									A072201	High Impedance		
									A090801	Signal Artifact/Noise		
							F26	No Health Consequences or Impact	A070909	Oversensing		
									A072201	High Impedance		
									A090801	Signal Artifact/Noise		
9525	2019	All other reportable incidents	37265	Dual-chamber implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A070909	Oversensing		
									A0712	Pacing Problem		
									A072201	High Impedance		
									A090801	Signal Artifact/Noise		
							F26	No Health Consequences or Impact	A070909	Oversensing		
									A0712	Pacing Problem		
									A072201	High Impedance		
									A090801	Signal Artifact/Noise		
9526	2020	All other reportable incidents	35852	Single-chamber implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A071302	Inappropriate/Inadequate Shock/Stimulation		
							F26	No Health Consequences or Impact	A071302	Inappropriate/Inadequate Shock/Stimulation		
9527	2020	All other reportable incidents	36241	Pacemaker/defibrillator lead	CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070909	Oversensing		
							F26	No Health Consequences or Impact	A070909	Oversensing		
9528	2020	Death	17882	Physiologic-monitoring defibrillation system	CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A071301	Failure to Deliver Shock/Stimulation		
							F02	Death	A071301	Failure to Deliver Shock/Stimulation		
							F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A071301	Failure to Deliver Shock/Stimulation		
9529	2019	All other reportable incidents	17882	Physiologic-monitoring defibrillation system	CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A071301	Failure to Deliver Shock/Stimulation		
9530	2019	All other reportable incidents	37265	Dual-chamber implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A070909	Oversensing		
9531	2019	All other reportable incidents	36241	Pacemaker/defibrillator lead	CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A072201	High Impedance		
9532	2019	All other reportable incidents	37265	Dual-chamber implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	07/08/2019	
9533	2020	All other reportable incidents	35223	Endocardial/interventricular septal pacing lead	CT1684	Cardiac pacemakers and associated devices	E060104	Bradycardia	A05	Mechanical Problem	02/07/2020	02/07/2020
									A070909	Oversensing	02/07/2020	02/07/2020
									A072202	Low Impedance	02/07/2020	02/07/2020
									A090801	Signal Artifact/Noise	02/07/2020	02/07/2020
							E2328	Obstruction/Occlusion	A05	Mechanical Problem	02/07/2020	02/07/2020
									A070909	Oversensing	02/07/2020	02/07/2020
									A072202	Low Impedance	02/07/2020	02/07/2020
									A090801	Signal Artifact/Noise	02/07/2020	02/07/2020
							F1905	Device Revision or Replacement	A05	Mechanical Problem	02/07/2020	02/07/2020
									A070909	Oversensing	02/07/2020	02/07/2020
									A072202	Low Impedance	02/07/2020	02/07/2020
									A090801	Signal Artifact/Noise	02/07/2020	02/07/2020
							F21	Unexpected Deterioration	A05	Mechanical Problem	02/07/2020	02/07/2020
									A070909	Oversensing	02/07/2020	02/07/2020
									A072202	Low Impedance	02/07/2020	02/07/2020
									A090801	Signal Artifact/Noise	02/07/2020	02/07/2020
9534	2019	All other reportable incidents	47267	Single-chamber implantable pacemaker, rate-responsive	CT141	Electrodes/Leads and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
					CT145	Cardiovascular devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
					CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
					CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
					CT233	Surgical	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		

					CT2406	Implantable	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
					CT2407	Active implantable	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
					CT2727	Active implant devices/ancillaries	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
					CT2750	Implantable cardiac electrical stimulators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
					CT2985	Active	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
					CT3001	Implantable metal	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
					CT3002	Implantable synthetic polymer	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
					CT3006	Implantable inorganic material	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
					CT314	Electrically powered	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
					CT321	Long-term surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
					CT449	Generators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
					CT981	Single use	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
					CT983	Surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
9535	2019	All other reportable incidents	36241	Pacemaker/defibrillator lead	CT1684	Cardiac pacemakers and associated devices	F1905	Device Revision or Replacement	A070910	Undersensing		
9536	2020	All other reportable incidents	36241	Pacemaker/defibrillator lead	CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A072201	High Impedance		
							F26	No Health Consequences or Impact	A072201	High Impedance		
9537	2020	All other reportable incidents	35852	Single-chamber implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A072201	High Impedance		
							F26	No Health Consequences or Impact	A072201	High Impedance		
9538	2020	All other reportable incidents	35852	Single-chamber implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	24/07/2020	24/07/2020
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	24/07/2020	24/07/2020
							F12	Serious Injury/ Illness/ Impairment	A070504	Premature Discharge of Battery	24/07/2020	24/07/2020
							F1903	Device Explantation	A070504	Premature Discharge of Battery	24/07/2020	24/07/2020
9539	2019	All other reportable incidents	37265	Dual-chamber implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E0112	Dizziness	A070504	Premature Discharge of Battery	08/08/2019	08/08/2019
									A071102	Failure to Interrogate	08/08/2019	08/08/2019
							E2312	Fatigue	A070504	Premature Discharge of Battery	08/08/2019	08/08/2019
									A071102	Failure to Interrogate	08/08/2019	08/08/2019
9540	2019	All other reportable incidents	17882	Physiologic-monitoring defibrillation system	CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A0719	Unexpected Shutdown		
							F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A0719	Unexpected Shutdown		
							F26	No Health Consequences or Impact	A0719	Unexpected Shutdown		
9541	2019	All other reportable incidents	36241	Pacemaker/defibrillator lead	CT1684	Cardiac pacemakers and associated devices	E233001	Chest Pain	A070102	High Capture Threshold	26/07/2019	26/07/2019
							F1905	Device Revision or Replacement	A070102	High Capture Threshold	26/07/2019	26/07/2019
9542	2019	All other reportable incidents	35852	Single-chamber implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A070909	Oversensing		
9543	2020	All other reportable incidents	35223	Endocardial/interventricular septal pacing lead	CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A0401	Break		
									A070909	Oversensing		
									A0712	Pacing Problem		
							F1901	Additional Surgery	A0401	Break		
									A070909	Oversensing		
									A0712	Pacing Problem		
9544	2020	Death	37265	Dual-chamber implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E011903	Syncope/Fainting	A1107	Data Problem	10/07/2020	10/07/2020
									A24	No Apparent Device Problem	10/07/2020	10/07/2020
							E2007	Fall	A1107	Data Problem	10/07/2020	10/07/2020
									A24	No Apparent Device Problem	10/07/2020	10/07/2020
							E2401	Insufficient Information	A1107	Data Problem	10/07/2020	10/07/2020
									A24	No Apparent Device Problem	10/07/2020	10/07/2020
							F02	Death	A1107	Data Problem	10/07/2020	10/07/2020
									A24	No Apparent Device Problem	10/07/2020	10/07/2020
9545	2020	All other reportable incidents	36241	Pacemaker/defibrillator lead	CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A072201	High Impedance		
							F26	No Health Consequences or Impact	A072201	High Impedance		
9546	2020	All other reportable incidents	36241	Pacemaker/defibrillator lead	CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A0722	Impedance Problem		
							F26	No Health Consequences or Impact	A0722	Impedance Problem		
9547	2020	All other reportable incidents	47270	Cardiac resynchronization therapy implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A0712	Pacing Problem		
							F24	Insufficient Health Impact Information	A0712	Pacing Problem		
9548	2020	All other reportable incidents	47265	Dual-chamber implantable pacemaker, rate-responsive	CT141	Electrodes/Leads and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	01/07/2020	17/09/2020
							F19	Surgical Intervention	A070504	Premature Discharge of Battery	01/07/2020	17/09/2020
					CT145	Cardiovascular devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	01/07/2020	17/09/2020
							F19	Surgical Intervention	A070504	Premature Discharge of Battery	01/07/2020	17/09/2020
					CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	01/07/2020	17/09/2020
							F19	Surgical Intervention	A070504	Premature Discharge of Battery	01/07/2020	17/09/2020
					CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	01/07/2020	17/09/2020
							F19	Surgical Intervention	A070504	Premature Discharge of Battery	01/07/2020	17/09/2020
					CT233	Surgical	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	01/07/2020	17/09/2020
							F19	Surgical Intervention	A070504	Premature Discharge of Battery	01/07/2020	17/09/2020
					CT2406	Implantable	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	01/07/2020	17/09/2020
							F19	Surgical Intervention	A070504	Premature Discharge of Battery	01/07/2020	17/09/2020
					CT2407	Active implantable	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	01/07/2020	17/09/2020
							F19	Surgical Intervention	A070504	Premature Discharge of Battery	01/07/2020	17/09/2020
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	01/07/2020	17/09/2020
							F19	Surgical Intervention	A070504	Premature Discharge of Battery	01/07/2020	17/09/2020
					CT2727	Active implant devices/ancillaries	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	01/07/2020	17/09/2020
							F19	Surgical Intervention	A070504	Premature Discharge of Battery	01/07/2020	17/09/2020
					CT2750	Implantable cardiac electrical stimulators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	01/07/2020	17/09/2020
							F19	Surgical Intervention	A070504	Premature Discharge of Battery	01/07/2020	17/09/2020
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	01/07/2020	17/09/2020
							F19	Surgical Intervention	A070504	Premature Discharge of Battery	01/07/2020	17/09/2020
					CT2985	Active	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	01/07/2020	17/09/2020
							F19	Surgical Intervention	A070504	Premature Discharge of Battery	01/07/2020	17/09/2020
					CT3001	Implantable metal	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	01/07/2020	17/09/2020
							F19	Surgical Intervention	A070504	Premature Discharge of Battery	01/07/2020	17/09/2020
					CT3002	Implantable synthetic polymer	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	01/07/2020	17/09/2020
							F19	Surgical Intervention	A070504	Premature Discharge of Battery	01/07/2020	17/09/2020
					CT3006	Implantable inorganic material	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	01/07/2020	17/09/2020
							F19	Surgical Intervention	A070504	Premature Discharge of Battery	01/07/2020	17/09/2020
					CT314	Electrically powered	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	01/07/2020	17/09/2020
							F19	Surgical Intervention	A070504	Premature Discharge of Battery	01/07/2020	17/09/2020
					CT321	Long-term surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	01/07/2020	17/09/2020

							F19	Surgical Intervention	A070504	Premature Discharge of Battery	01/07/2020	17/09/2020
					CT449	Generators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	01/07/2020	17/09/2020
							F19	Surgical Intervention	A070504	Premature Discharge of Battery	01/07/2020	17/09/2020
					CT981	Single use	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	01/07/2020	17/09/2020
							F19	Surgical Intervention	A070504	Premature Discharge of Battery	01/07/2020	17/09/2020
					CT983	Surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	01/07/2020	17/09/2020
							F19	Surgical Intervention	A070504	Premature Discharge of Battery	01/07/2020	17/09/2020
9549	2019	All other reportable incidents	37265	Dual-chamber implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	31/08/2019	
9550	2020	All other reportable incidents	17882	Physiologic-monitoring defibrillation system	CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070803	Failure to Power Up		
							F27	Problem Identified During Non-Clinical Procedure	A070803	Failure to Power Up		
9551	2020	All other reportable incidents	17882	Physiologic-monitoring defibrillation system	CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070803	Failure to Power Up		
							F27	Problem Identified During Non-Clinical Procedure	A070803	Failure to Power Up		
9552	2019	All other reportable incidents	37265	Dual-chamber implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A072201	High Impedance		
9553	2019	All other reportable incidents	35852	Single-chamber implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A070909	Oversensing		
9554	2019	All other reportable incidents	47270	Cardiac resynchronization therapy implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A071302	Inappropriate/Inadequate Shock/Stimulation		
9555	2020	All other reportable incidents	35852	Single-chamber implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E2104	Electric Shock	A071302	Inappropriate/Inadequate Shock/Stimulation		
							F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A071302	Inappropriate/Inadequate Shock/Stimulation		
							F26	No Health Consequences or Impact	A071302	Inappropriate/Inadequate Shock/Stimulation		
9556	2019	All other reportable incidents	35852	Single-chamber implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A071302	Inappropriate/Inadequate Shock/Stimulation		
9557	2019	All other reportable incidents	47265	Dual-chamber implantable pacemaker, rate-responsive	CT141	Electrodes/Leads and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070909	Oversensing		
					CT145	Cardiovascular devices	E2403	No Clinical Signs, Symptoms or Conditions	A070909	Oversensing		
					CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070909	Oversensing		
					CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A070909	Oversensing		
					CT233	Surgical	E2403	No Clinical Signs, Symptoms or Conditions	A070909	Oversensing		
					CT2406	Implantable	E2403	No Clinical Signs, Symptoms or Conditions	A070909	Oversensing		
					CT2407	Active implantable	E2403	No Clinical Signs, Symptoms or Conditions	A070909	Oversensing		
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A070909	Oversensing		
					CT2727	Active implant devices/ancillaries	E2403	No Clinical Signs, Symptoms or Conditions	A070909	Oversensing		
					CT2750	Implantable cardiac electrical stimulators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070909	Oversensing		
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A070909	Oversensing		
					CT2985	Active	E2403	No Clinical Signs, Symptoms or Conditions	A070909	Oversensing		
					CT3001	Implantable metal	E2403	No Clinical Signs, Symptoms or Conditions	A070909	Oversensing		
					CT3002	Implantable synthetic polymer	E2403	No Clinical Signs, Symptoms or Conditions	A070909	Oversensing		
					CT3006	Implantable inorganic material	E2403	No Clinical Signs, Symptoms or Conditions	A070909	Oversensing		
					CT314	Electrically powered	E2403	No Clinical Signs, Symptoms or Conditions	A070909	Oversensing		
					CT321	Long-term surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A070909	Oversensing		
					CT449	Generators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070909	Oversensing		
					CT981	Single use	E2403	No Clinical Signs, Symptoms or Conditions	A070909	Oversensing		
					CT983	Surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A070909	Oversensing		
9558	2020	All other reportable incidents	35852	Single-chamber implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E2104	Electric Shock	A070909	Oversensing		
									A071302	Inappropriate/Inadequate Shock/Stimulation		
							E2403	No Clinical Signs, Symptoms or Conditions	A070909	Oversensing		
									A071302	Inappropriate/Inadequate Shock/Stimulation		
							F26	No Health Consequences or Impact	A070909	Oversensing		
									A071302	Inappropriate/Inadequate Shock/Stimulation		
9559	2020	All other reportable incidents	17882	Physiologic-monitoring defibrillation system	CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A07	Electrical /Electronic Property Problem		
							F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A07	Electrical /Electronic Property Problem		
							F26	No Health Consequences or Impact	A07	Electrical /Electronic Property Problem		
9560	2019	All other reportable incidents	36241	Pacemaker/defibrillator lead	CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A05	Mechanical Problem		
9561	2020	All other reportable incidents	47270	Cardiac resynchronization therapy implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A070909	Oversensing		
							F26	No Health Consequences or Impact	A070909	Oversensing		
9562	2020	All other reportable incidents	35852	Single-chamber implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem	13/12/2020	13/12/2020
									A070504	Premature Discharge of Battery	13/12/2020	13/12/2020
							F08	Hospitalization or Prolonged Hospitalization	A0705	Battery Problem	13/12/2020	13/12/2020
									A070504	Premature Discharge of Battery	13/12/2020	13/12/2020
							F12	Serious Injury/ Illness/ Impairment	A0705	Battery Problem	13/12/2020	13/12/2020
									A070504	Premature Discharge of Battery	13/12/2020	13/12/2020
							F19	Surgical Intervention	A0705	Battery Problem	13/12/2020	13/12/2020
									A070504	Premature Discharge of Battery	13/12/2020	13/12/2020
9563	2020	All other reportable incidents	17882	Physiologic-monitoring defibrillation system	CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070908	Failure to Sense		
							F26	No Health Consequences or Impact	A070908	Failure to Sense		
9564	2019	All other reportable incidents	35852	Single-chamber implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A070909	Oversensing		
9565	2020	All other reportable incidents	37265	Dual-chamber implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	09/07/2020	09/07/2020
									A071102	Failure to Interrogate	09/07/2020	09/07/2020
							F1901	Additional Surgery	A070504	Premature Discharge of Battery	09/07/2020	09/07/2020
									A071102	Failure to Interrogate	09/07/2020	09/07/2020
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	09/07/2020	09/07/2020
									A071102	Failure to Interrogate	09/07/2020	09/07/2020
9566	2020	All other reportable incidents	37265	Dual-chamber implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	27/07/2020	27/07/2020
							F1901	Additional Surgery	A070504	Premature Discharge of Battery	27/07/2020	27/07/2020
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	27/07/2020	27/07/2020
9567	2020	All other reportable incidents	35223	Endocardial/interventricular septal pacing lead	CT1684	Cardiac pacemakers and associated devices	E011903	Syncope/Fainting	A0701	Capturing Problem		
									A070104	Unstable Capture Threshold		
							E060104	Bradycardia	A0701	Capturing Problem		
									A070104	Unstable Capture Threshold		
							F1203	Life Threatening Illness or Injury	A0701	Capturing Problem		
									A070104	Unstable Capture Threshold		
							F1905	Device Revision or Replacement	A0701	Capturing Problem		
									A070104	Unstable Capture Threshold		
9568	2020	All other reportable incidents	36241	Pacemaker/defibrillator lead	CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070909	Oversensing		
									A090801	Signal Artifact/Noise		
									A1301	Failure to Read Input Signal		
							F15	Recognized Device or Procedural Complication	A070909	Oversensing		
									A090801	Signal Artifact/Noise		
									A1301	Failure to Read Input Signal		
							F26	No Health Consequences or Impact	A070909	Oversensing		

									A090801	Signal Artifact/Noise		
									A1301	Failure to Read Input Signal		
9569	2020	All other reportable incidents	37265	Dual-chamber implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A072201	High Impedance		
							F26	No Health Consequences or Impact	A072201	High Impedance		
9570	2020	All other reportable incidents	36241	Pacemaker/defibrillator lead	CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A072201	High Impedance		
							F26	No Health Consequences or Impact	A072201	High Impedance		
9571	2020	All other reportable incidents	37265	Dual-chamber implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	08/07/2020	08/07/2020
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	08/07/2020	08/07/2020
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery	08/07/2020	08/07/2020
9572	2019	All other reportable incidents	35852	Single-chamber implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A070909	Oversensing		
9573	2019	All other reportable incidents	35852	Single-chamber implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A0203	Defective Device	01/08/2019	01/08/2019
9574	2020	All other reportable incidents	37265	Dual-chamber implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
9575	2019	All other reportable incidents	37265	Dual-chamber implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A070902	Increased Sensitivity		
9576	2020	All other reportable incidents	47265	Dual-chamber implantable pacemaker, rate-responsive	CT141	Electrodes/Leads and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	29/06/2020	29/06/2020
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	29/06/2020	29/06/2020
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery	29/06/2020	29/06/2020
					CT145	Cardiovascular devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	29/06/2020	29/06/2020
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	29/06/2020	29/06/2020
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery	29/06/2020	29/06/2020
					CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	29/06/2020	29/06/2020
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	29/06/2020	29/06/2020
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery	29/06/2020	29/06/2020
					CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	29/06/2020	29/06/2020
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	29/06/2020	29/06/2020
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery	29/06/2020	29/06/2020
					CT233	Surgical	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	29/06/2020	29/06/2020
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	29/06/2020	29/06/2020
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery	29/06/2020	29/06/2020
					CT2406	Implantable	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	29/06/2020	29/06/2020
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	29/06/2020	29/06/2020
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery	29/06/2020	29/06/2020
					CT2407	Active implantable	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	29/06/2020	29/06/2020
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	29/06/2020	29/06/2020
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery	29/06/2020	29/06/2020
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	29/06/2020	29/06/2020
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	29/06/2020	29/06/2020
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery	29/06/2020	29/06/2020
					CT2727	Active implant devices/ancillaries	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	29/06/2020	29/06/2020
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	29/06/2020	29/06/2020
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery	29/06/2020	29/06/2020
					CT2750	Implantable cardiac electrical stimulators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	29/06/2020	29/06/2020
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	29/06/2020	29/06/2020
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery	29/06/2020	29/06/2020
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	29/06/2020	29/06/2020
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	29/06/2020	29/06/2020
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery	29/06/2020	29/06/2020
					CT2985	Active	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	29/06/2020	29/06/2020
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	29/06/2020	29/06/2020
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery	29/06/2020	29/06/2020
					CT3001	Implantable metal	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	29/06/2020	29/06/2020
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	29/06/2020	29/06/2020
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery	29/06/2020	29/06/2020
					CT3002	Implantable synthetic polymer	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	29/06/2020	29/06/2020
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	29/06/2020	29/06/2020
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery	29/06/2020	29/06/2020
					CT3006	Implantable inorganic material	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	29/06/2020	29/06/2020
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	29/06/2020	29/06/2020
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery	29/06/2020	29/06/2020
					CT314	Electrically powered	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	29/06/2020	29/06/2020
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	29/06/2020	29/06/2020
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery	29/06/2020	29/06/2020
					CT321	Long-term surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	29/06/2020	29/06/2020
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	29/06/2020	29/06/2020
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery	29/06/2020	29/06/2020
					CT449	Generators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	29/06/2020	29/06/2020
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	29/06/2020	29/06/2020
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery	29/06/2020	29/06/2020
					CT981	Single use	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	29/06/2020	29/06/2020
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	29/06/2020	29/06/2020
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery	29/06/2020	29/06/2020
					CT983	Surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	29/06/2020	29/06/2020
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	29/06/2020	29/06/2020
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery	29/06/2020	29/06/2020
9577	2019	All other reportable incidents	36241	Pacemaker/defibrillator lead	CT1684	Cardiac pacemakers and associated devices	F1905	Device Revision or Replacement	A010402	Migration		
9578	2020	All other reportable incidents	35223	Endocardial/interventricular septal pacing lead	CT1684	Cardiac pacemakers and associated devices	E0112	Dizziness	A0701	Capturing Problem	06/07/2020	06/07/2020
									A070103	Intermittent Capture	06/07/2020	06/07/2020
							F11	Non-Serious Injury/ Illness/ Impairment	A0701	Capturing Problem	06/07/2020	06/07/2020
									A070103	Intermittent Capture	06/07/2020	06/07/2020
							F1905	Device Revision or Replacement	A0701	Capturing Problem	06/07/2020	06/07/2020
									A070103	Intermittent Capture	06/07/2020	06/07/2020
9579	2020	All other reportable incidents	36241	Pacemaker/defibrillator lead	CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A072201	High Impedance		
							F26	No Health Consequences or Impact	A072201	High Impedance		
9580	2020	All other reportable incidents	35852	Single-chamber implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A070909	Oversensing		
									A070910	Undersensing		
							F26	No Health Consequences or Impact	A070909	Oversensing		
									A070910	Undersensing		

9581	2019	All other reportable incidents	36241	Pacemaker/defibrillator lead	CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A010402	Migration		
9582	2019	All other reportable incidents	36241	Pacemaker/defibrillator lead	CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070909	Oversensing		
9583	2019	All other reportable incidents	37265	Dual-chamber implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A070909	Oversensing	01/08/2019	
9584	2019	All other reportable incidents	35852	Single-chamber implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A070909	Oversensing		
9585	2019	All other reportable incidents	35223	Endocardial/interventricular septal pacing lead	CT1684	Cardiac pacemakers and associated devices	F1905	Device Revision or Replacement	A0405	Degraded	06/08/2019	
9586	2019	All other reportable incidents	37265	Dual-chamber implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	21/08/2019	21/08/2019
9587	2019	All other reportable incidents	35223	Endocardial/interventricular septal pacing lead	CT1684	Cardiac pacemakers and associated devices	F1905	Device Revision or Replacement	A070909	Oversensing		
9588	2019	All other reportable incidents	35852	Single-chamber implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	F1905	Device Revision or Replacement	A0705	Battery Problem	30/08/2019	
9589	2019	All other reportable incidents	17882	Physiologic-monitoring defibrillation system	CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A	Annex A		
9590	2020	All other reportable incidents	47267	Single-chamber implantable pacemaker, rate-responsive	CT141	Electrodes/Leads and associated devices	E060104	Bradycardia	A1107	Data Problem		
							F08	Hospitalization or Prolonged Hospitalization	A1107	Data Problem		
							F1905	Device Revision or Replacement	A1107	Data Problem		
					CT145	Cardiovascular devices	E060104	Bradycardia	A1107	Data Problem		
							F08	Hospitalization or Prolonged Hospitalization	A1107	Data Problem		
							F1905	Device Revision or Replacement	A1107	Data Problem		
					CT1684	Cardiac pacemakers and associated devices	E060104	Bradycardia	A1107	Data Problem		
							F08	Hospitalization or Prolonged Hospitalization	A1107	Data Problem		
							F1905	Device Revision or Replacement	A1107	Data Problem		
					CT1685	Cardiac pacemakers	E060104	Bradycardia	A1107	Data Problem		
							F08	Hospitalization or Prolonged Hospitalization	A1107	Data Problem		
							F1905	Device Revision or Replacement	A1107	Data Problem		
					CT233	Surgical	E060104	Bradycardia	A1107	Data Problem		
							F08	Hospitalization or Prolonged Hospitalization	A1107	Data Problem		
							F1905	Device Revision or Replacement	A1107	Data Problem		
					CT2406	Implantable	E060104	Bradycardia	A1107	Data Problem		
							F08	Hospitalization or Prolonged Hospitalization	A1107	Data Problem		
							F1905	Device Revision or Replacement	A1107	Data Problem		
					CT2407	Active implantable	E060104	Bradycardia	A1107	Data Problem		
							F08	Hospitalization or Prolonged Hospitalization	A1107	Data Problem		
							F1905	Device Revision or Replacement	A1107	Data Problem		
					CT257	Implantable cardiac pacemakers	E060104	Bradycardia	A1107	Data Problem		
							F08	Hospitalization or Prolonged Hospitalization	A1107	Data Problem		
							F1905	Device Revision or Replacement	A1107	Data Problem		
					CT2727	Active implant devices/ancillaries	E060104	Bradycardia	A1107	Data Problem		
							F08	Hospitalization or Prolonged Hospitalization	A1107	Data Problem		
							F1905	Device Revision or Replacement	A1107	Data Problem		
					CT2750	Implantable cardiac electrical stimulators and associated devices	E060104	Bradycardia	A1107	Data Problem		
							F08	Hospitalization or Prolonged Hospitalization	A1107	Data Problem		
							F1905	Device Revision or Replacement	A1107	Data Problem		
					CT2866	Implantable cardiac electrical stimulators	E060104	Bradycardia	A1107	Data Problem		
							F08	Hospitalization or Prolonged Hospitalization	A1107	Data Problem		
							F1905	Device Revision or Replacement	A1107	Data Problem		
					CT2985	Active	E060104	Bradycardia	A1107	Data Problem		
							F08	Hospitalization or Prolonged Hospitalization	A1107	Data Problem		
							F1905	Device Revision or Replacement	A1107	Data Problem		
					CT3001	Implantable metal	E060104	Bradycardia	A1107	Data Problem		
							F08	Hospitalization or Prolonged Hospitalization	A1107	Data Problem		
							F1905	Device Revision or Replacement	A1107	Data Problem		
					CT3002	Implantable synthetic polymer	E060104	Bradycardia	A1107	Data Problem		
							F08	Hospitalization or Prolonged Hospitalization	A1107	Data Problem		
							F1905	Device Revision or Replacement	A1107	Data Problem		
					CT3006	Implantable inorganic material	E060104	Bradycardia	A1107	Data Problem		
							F08	Hospitalization or Prolonged Hospitalization	A1107	Data Problem		
							F1905	Device Revision or Replacement	A1107	Data Problem		
					CT314	Electrically powered	E060104	Bradycardia	A1107	Data Problem		
							F08	Hospitalization or Prolonged Hospitalization	A1107	Data Problem		
							F1905	Device Revision or Replacement	A1107	Data Problem		
					CT321	Long-term surgical invasive	E060104	Bradycardia	A1107	Data Problem		
							F08	Hospitalization or Prolonged Hospitalization	A1107	Data Problem		
							F1905	Device Revision or Replacement	A1107	Data Problem		
					CT449	Generators and associated devices	E060104	Bradycardia	A1107	Data Problem		
							F08	Hospitalization or Prolonged Hospitalization	A1107	Data Problem		
							F1905	Device Revision or Replacement	A1107	Data Problem		
					CT981	Single use	E060104	Bradycardia	A1107	Data Problem		
							F08	Hospitalization or Prolonged Hospitalization	A1107	Data Problem		
							F1905	Device Revision or Replacement	A1107	Data Problem		
					CT983	Surgical invasive	E060104	Bradycardia	A1107	Data Problem		
							F08	Hospitalization or Prolonged Hospitalization	A1107	Data Problem		
							F1905	Device Revision or Replacement	A1107	Data Problem		
9591	2020	All other reportable incidents	47270	Cardiac resynchronization therapy implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	25/08/2020	25/08/2020
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	25/08/2020	25/08/2020
							F12	Serious Injury/ Illness/ Impairment	A070504	Premature Discharge of Battery	25/08/2020	25/08/2020
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	25/08/2020	25/08/2020
9592	2020	All other reportable incidents	35852	Single-chamber implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E2104	Electric Shock	A070909	Oversensing		
									A071302	Inappropriate/Inadequate Shock/Stimulation		
							F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A070909	Oversensing		
									A071302	Inappropriate/Inadequate Shock/Stimulation		
							F26	No Health Consequences or Impact	A070909	Oversensing		
									A071302	Inappropriate/Inadequate Shock/Stimulation		
9593	2019	All other reportable incidents	35852	Single-chamber implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A070909	Oversensing		
9594	2019	All other reportable incidents	35852	Single-chamber implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A	Annex A		
9595	2020	All other reportable incidents	47270	Cardiac resynchronization therapy implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	15/07/2020	15/07/2020
							F1901	Additional Surgery	A070504	Premature Discharge of Battery	15/07/2020	15/07/2020
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	15/07/2020	15/07/2020
9596	2020	All other reportable incidents	47263	Cardiac resynchronization therapy implantable pacemaker	CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	23/06/2020	23/06/2020
							F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A070504	Premature Discharge of Battery	23/06/2020	23/06/2020

							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	23/06/2020	23/06/2020
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery	23/06/2020	23/06/2020
					CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	23/06/2020	23/06/2020
							F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A070504	Premature Discharge of Battery	23/06/2020	23/06/2020
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	23/06/2020	23/06/2020
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery	23/06/2020	23/06/2020
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	23/06/2020	23/06/2020
							F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A070504	Premature Discharge of Battery	23/06/2020	23/06/2020
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	23/06/2020	23/06/2020
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery	23/06/2020	23/06/2020
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	23/06/2020	23/06/2020
							F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A070504	Premature Discharge of Battery	23/06/2020	23/06/2020
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	23/06/2020	23/06/2020
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery	23/06/2020	23/06/2020
9597	2020	All other reportable incidents	47265	Dual-chamber implantable pacemaker, rate-responsive	CT141	Electrodes/Leads and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A072201	High Impedance		
							F26	No Health Consequences or Impact	A072201	High Impedance		
					CT145	Cardiovascular devices	E2403	No Clinical Signs, Symptoms or Conditions	A072201	High Impedance		
							F26	No Health Consequences or Impact	A072201	High Impedance		
					CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A072201	High Impedance		
							F26	No Health Consequences or Impact	A072201	High Impedance		
					CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A072201	High Impedance		
							F26	No Health Consequences or Impact	A072201	High Impedance		
					CT233	Surgical	E2403	No Clinical Signs, Symptoms or Conditions	A072201	High Impedance		
							F26	No Health Consequences or Impact	A072201	High Impedance		
					CT2406	Implantable	E2403	No Clinical Signs, Symptoms or Conditions	A072201	High Impedance		
							F26	No Health Consequences or Impact	A072201	High Impedance		
					CT2407	Active implantable	E2403	No Clinical Signs, Symptoms or Conditions	A072201	High Impedance		
							F26	No Health Consequences or Impact	A072201	High Impedance		
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A072201	High Impedance		
							F26	No Health Consequences or Impact	A072201	High Impedance		
					CT2727	Active implant devices/ancillaries	E2403	No Clinical Signs, Symptoms or Conditions	A072201	High Impedance		
							F26	No Health Consequences or Impact	A072201	High Impedance		
					CT2750	Implantable cardiac electrical stimulators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A072201	High Impedance		
							F26	No Health Consequences or Impact	A072201	High Impedance		
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A072201	High Impedance		
							F26	No Health Consequences or Impact	A072201	High Impedance		
					CT2985	Active	E2403	No Clinical Signs, Symptoms or Conditions	A072201	High Impedance		
							F26	No Health Consequences or Impact	A072201	High Impedance		
					CT3001	Implantable metal	E2403	No Clinical Signs, Symptoms or Conditions	A072201	High Impedance		
							F26	No Health Consequences or Impact	A072201	High Impedance		
					CT3002	Implantable synthetic polymer	E2403	No Clinical Signs, Symptoms or Conditions	A072201	High Impedance		
							F26	No Health Consequences or Impact	A072201	High Impedance		
					CT3006	Implantable inorganic material	E2403	No Clinical Signs, Symptoms or Conditions	A072201	High Impedance		
							F26	No Health Consequences or Impact	A072201	High Impedance		
					CT314	Electrically powered	E2403	No Clinical Signs, Symptoms or Conditions	A072201	High Impedance		
							F26	No Health Consequences or Impact	A072201	High Impedance		
					CT321	Long-term surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A072201	High Impedance		
							F26	No Health Consequences or Impact	A072201	High Impedance		
					CT449	Generators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A072201	High Impedance		
							F26	No Health Consequences or Impact	A072201	High Impedance		
					CT981	Single use	E2403	No Clinical Signs, Symptoms or Conditions	A072201	High Impedance		
							F26	No Health Consequences or Impact	A072201	High Impedance		
					CT983	Surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A072201	High Impedance		
							F26	No Health Consequences or Impact	A072201	High Impedance		
9598	2020	All other reportable incidents	47267	Single-chamber implantable pacemaker, rate-responsive	CT141	Electrodes/Leads and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A070504	Premature Discharge of Battery		
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT145	Cardiovascular devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A070504	Premature Discharge of Battery		
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A070504	Premature Discharge of Battery		
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A070504	Premature Discharge of Battery		
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT233	Surgical	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A070504	Premature Discharge of Battery		
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT2406	Implantable	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A070504	Premature Discharge of Battery		
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT2407	Active implantable	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A070504	Premature Discharge of Battery		
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A070504	Premature Discharge of Battery		
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT2727	Active implant devices/ancillaries	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A070504	Premature Discharge of Battery		

							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT2750	Implantable cardiac electrical stimulators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A070504	Premature Discharge of Battery		
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A070504	Premature Discharge of Battery		
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT2985	Active	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A070504	Premature Discharge of Battery		
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT3001	Implantable metal	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A070504	Premature Discharge of Battery		
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT3002	Implantable synthetic polymer	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A070504	Premature Discharge of Battery		
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT3006	Implantable inorganic material	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A070504	Premature Discharge of Battery		
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT314	Electrically powered	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A070504	Premature Discharge of Battery		
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT321	Long-term surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A070504	Premature Discharge of Battery		
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT449	Generators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A070504	Premature Discharge of Battery		
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT981	Single use	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A070504	Premature Discharge of Battery		
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT983	Surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A070504	Premature Discharge of Battery		
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
9599	2020	All other reportable incidents	47265	Dual-chamber implantable pacemaker, rate-responsive	CT141	Electrodes/Leads and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	25/08/2020	25/08/2020
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	25/08/2020	25/08/2020
							F12	Serious Injury/ Illness/ Impairment	A070504	Premature Discharge of Battery	25/08/2020	25/08/2020
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	25/08/2020	25/08/2020
					CT145	Cardiovascular devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	25/08/2020	25/08/2020
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	25/08/2020	25/08/2020
							F12	Serious Injury/ Illness/ Impairment	A070504	Premature Discharge of Battery	25/08/2020	25/08/2020
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	25/08/2020	25/08/2020
					CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	25/08/2020	25/08/2020
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	25/08/2020	25/08/2020
							F12	Serious Injury/ Illness/ Impairment	A070504	Premature Discharge of Battery	25/08/2020	25/08/2020
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	25/08/2020	25/08/2020
					CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	25/08/2020	25/08/2020
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	25/08/2020	25/08/2020
							F12	Serious Injury/ Illness/ Impairment	A070504	Premature Discharge of Battery	25/08/2020	25/08/2020
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	25/08/2020	25/08/2020
					CT233	Surgical	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	25/08/2020	25/08/2020
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	25/08/2020	25/08/2020
							F12	Serious Injury/ Illness/ Impairment	A070504	Premature Discharge of Battery	25/08/2020	25/08/2020
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	25/08/2020	25/08/2020
					CT2406	Implantable	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	25/08/2020	25/08/2020
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	25/08/2020	25/08/2020
							F12	Serious Injury/ Illness/ Impairment	A070504	Premature Discharge of Battery	25/08/2020	25/08/2020
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	25/08/2020	25/08/2020
					CT2407	Active implantable	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	25/08/2020	25/08/2020
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	25/08/2020	25/08/2020
							F12	Serious Injury/ Illness/ Impairment	A070504	Premature Discharge of Battery	25/08/2020	25/08/2020
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	25/08/2020	25/08/2020
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	25/08/2020	25/08/2020
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	25/08/2020	25/08/2020
							F12	Serious Injury/ Illness/ Impairment	A070504	Premature Discharge of Battery	25/08/2020	25/08/2020
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	25/08/2020	25/08/2020
					CT2727	Active implant devices/ancillaries	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	25/08/2020	25/08/2020
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	25/08/2020	25/08/2020
							F12	Serious Injury/ Illness/ Impairment	A070504	Premature Discharge of Battery	25/08/2020	25/08/2020
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	25/08/2020	25/08/2020
					CT2750	Implantable cardiac electrical stimulators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	25/08/2020	25/08/2020
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	25/08/2020	25/08/2020
							F12	Serious Injury/ Illness/ Impairment	A070504	Premature Discharge of Battery	25/08/2020	25/08/2020
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	25/08/2020	25/08/2020
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	25/08/2020	25/08/2020
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	25/08/2020	25/08/2020

							F12	Serious Injury/ Illness/ Impairment	A070504	Premature Discharge of Battery	25/08/2020	25/08/2020
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	25/08/2020	25/08/2020
					CT2985	Active	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	25/08/2020	25/08/2020
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	25/08/2020	25/08/2020
							F12	Serious Injury/ Illness/ Impairment	A070504	Premature Discharge of Battery	25/08/2020	25/08/2020
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	25/08/2020	25/08/2020
					CT3001	Implantable metal	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	25/08/2020	25/08/2020
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	25/08/2020	25/08/2020
							F12	Serious Injury/ Illness/ Impairment	A070504	Premature Discharge of Battery	25/08/2020	25/08/2020
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	25/08/2020	25/08/2020
					CT3002	Implantable synthetic polymer	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	25/08/2020	25/08/2020
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	25/08/2020	25/08/2020
							F12	Serious Injury/ Illness/ Impairment	A070504	Premature Discharge of Battery	25/08/2020	25/08/2020
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	25/08/2020	25/08/2020
					CT3006	Implantable inorganic material	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	25/08/2020	25/08/2020
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	25/08/2020	25/08/2020
							F12	Serious Injury/ Illness/ Impairment	A070504	Premature Discharge of Battery	25/08/2020	25/08/2020
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	25/08/2020	25/08/2020
					CT314	Electrically powered	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	25/08/2020	25/08/2020
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	25/08/2020	25/08/2020
							F12	Serious Injury/ Illness/ Impairment	A070504	Premature Discharge of Battery	25/08/2020	25/08/2020
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	25/08/2020	25/08/2020
					CT321	Long-term surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	25/08/2020	25/08/2020
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	25/08/2020	25/08/2020
							F12	Serious Injury/ Illness/ Impairment	A070504	Premature Discharge of Battery	25/08/2020	25/08/2020
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	25/08/2020	25/08/2020
					CT449	Generators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	25/08/2020	25/08/2020
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	25/08/2020	25/08/2020
							F12	Serious Injury/ Illness/ Impairment	A070504	Premature Discharge of Battery	25/08/2020	25/08/2020
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	25/08/2020	25/08/2020
					CT981	Single use	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	25/08/2020	25/08/2020
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	25/08/2020	25/08/2020
							F12	Serious Injury/ Illness/ Impairment	A070504	Premature Discharge of Battery	25/08/2020	25/08/2020
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	25/08/2020	25/08/2020
					CT983	Surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	25/08/2020	25/08/2020
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	25/08/2020	25/08/2020
							F12	Serious Injury/ Illness/ Impairment	A070504	Premature Discharge of Battery	25/08/2020	25/08/2020
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	25/08/2020	25/08/2020
9600	2020	All other reportable incidents	47267	Single-chamber implantable pacemaker, rate-responsive	CT141	Electrodes/Leads and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT145	Cardiovascular devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT233	Surgical	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT2406	Implantable	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT2407	Active implantable	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT2727	Active implant devices/ancillaries	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT2750	Implantable cardiac electrical stimulators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT2985	Active	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT3001	Implantable metal	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT3002	Implantable synthetic polymer	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT3006	Implantable inorganic material	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT314	Electrically powered	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT321	Long-term surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A070504	Premature Discharge of Battery		

							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT449	Generators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT981	Single use	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT983	Surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
9601	2020	All other reportable incidents	37265	Dual-chamber implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
									A1601	Device Alarm System		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
									A1601	Device Alarm System		
9602	2019	All other reportable incidents	47265	Dual-chamber implantable pacemaker, rate-responsive	CT141	Electrodes/Leads and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	11/12/2019	11/12/2019
					CT145	Cardiovascular devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	11/12/2019	11/12/2019
					CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	11/12/2019	11/12/2019
					CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	11/12/2019	11/12/2019
					CT233	Surgical	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	11/12/2019	11/12/2019
					CT2406	Implantable	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	11/12/2019	11/12/2019
					CT2407	Active implantable	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	11/12/2019	11/12/2019
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	11/12/2019	11/12/2019
					CT2727	Active implant devices/ancillaries	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	11/12/2019	11/12/2019
					CT2750	Implantable cardiac electrical stimulators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	11/12/2019	11/12/2019
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	11/12/2019	11/12/2019
					CT2985	Active	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	11/12/2019	11/12/2019
					CT3001	Implantable metal	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	11/12/2019	11/12/2019
					CT3002	Implantable synthetic polymer	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	11/12/2019	11/12/2019
					CT3006	Implantable inorganic material	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	11/12/2019	11/12/2019
					CT314	Electrically powered	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	11/12/2019	11/12/2019
					CT321	Long-term surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	11/12/2019	11/12/2019
					CT449	Generators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	11/12/2019	11/12/2019
					CT981	Single use	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	11/12/2019	11/12/2019
					CT983	Surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	11/12/2019	11/12/2019
9603	2019	All other reportable incidents	36241	Pacemaker/defibrillator lead	CT1684	Cardiac pacemakers and associated devices	F1905	Device Revision or Replacement	A	Annex A		
9604	2020	All other reportable incidents	35852	Single-chamber implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E2104	Electric Shock	A070909	Oversensing		
									A071302	Inappropriate/Inadequate Shock/Stimulation		
							E2403	No Clinical Signs, Symptoms or Conditions	A070909	Oversensing		
									A071302	Inappropriate/Inadequate Shock/Stimulation		
							F26	No Health Consequences or Impact	A070909	Oversensing		
									A071302	Inappropriate/Inadequate Shock/Stimulation		
9605	2020	All other reportable incidents	17882	Physiologic-monitoring defibrillation system	CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A0713	Defibrillation/Stimulation Problem		
							F27	Problem Identified During Non-Clinical Procedure	A0713	Defibrillation/Stimulation Problem		
9606	2020	All other reportable incidents	17882	Physiologic-monitoring defibrillation system	CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070802	Intermittent Loss of Power		
							F27	Problem Identified During Non-Clinical Procedure	A070802	Intermittent Loss of Power		
9607	2020	All other reportable incidents	47265	Dual-chamber implantable pacemaker, rate-responsive	CT141	Electrodes/Leads and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	10/08/2020	10/08/2020
							F19	Surgical Intervention	A070504	Premature Discharge of Battery	10/08/2020	10/08/2020
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	10/08/2020	10/08/2020
					CT145	Cardiovascular devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	10/08/2020	10/08/2020
							F19	Surgical Intervention	A070504	Premature Discharge of Battery	10/08/2020	10/08/2020
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	10/08/2020	10/08/2020
					CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	10/08/2020	10/08/2020
							F19	Surgical Intervention	A070504	Premature Discharge of Battery	10/08/2020	10/08/2020
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	10/08/2020	10/08/2020
					CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	10/08/2020	10/08/2020
							F19	Surgical Intervention	A070504	Premature Discharge of Battery	10/08/2020	10/08/2020
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	10/08/2020	10/08/2020
					CT233	Surgical	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	10/08/2020	10/08/2020
							F19	Surgical Intervention	A070504	Premature Discharge of Battery	10/08/2020	10/08/2020
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	10/08/2020	10/08/2020
					CT2406	Implantable	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	10/08/2020	10/08/2020
							F19	Surgical Intervention	A070504	Premature Discharge of Battery	10/08/2020	10/08/2020
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	10/08/2020	10/08/2020
					CT2407	Active implantable	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	10/08/2020	10/08/2020
							F19	Surgical Intervention	A070504	Premature Discharge of Battery	10/08/2020	10/08/2020
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	10/08/2020	10/08/2020
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	10/08/2020	10/08/2020
							F19	Surgical Intervention	A070504	Premature Discharge of Battery	10/08/2020	10/08/2020
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	10/08/2020	10/08/2020
					CT2727	Active implant devices/ancillaries	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	10/08/2020	10/08/2020
							F19	Surgical Intervention	A070504	Premature Discharge of Battery	10/08/2020	10/08/2020
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	10/08/2020	10/08/2020
					CT2750	Implantable cardiac electrical stimulators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	10/08/2020	10/08/2020
							F19	Surgical Intervention	A070504	Premature Discharge of Battery	10/08/2020	10/08/2020
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	10/08/2020	10/08/2020
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	10/08/2020	10/08/2020
							F19	Surgical Intervention	A070504	Premature Discharge of Battery	10/08/2020	10/08/2020
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	10/08/2020	10/08/2020
					CT2985	Active	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	10/08/2020	10/08/2020
							F19	Surgical Intervention	A070504	Premature Discharge of Battery	10/08/2020	10/08/2020
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	10/08/2020	10/08/2020
					CT3001	Implantable metal	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	10/08/2020	10/08/2020
							F19	Surgical Intervention	A070504	Premature Discharge of Battery	10/08/2020	10/08/2020
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	10/08/2020	10/08/2020
					CT3002	Implantable synthetic polymer	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	10/08/2020	10/08/2020
							F19	Surgical Intervention	A070504	Premature Discharge of Battery	10/08/2020	10/08/2020
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	10/08/2020	10/08/2020

					CT3006	Implantable inorganic material	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	10/08/2020	10/08/2020
							F19	Surgical Intervention	A070504	Premature Discharge of Battery	10/08/2020	10/08/2020
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	10/08/2020	10/08/2020
					CT314	Electrically powered	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	10/08/2020	10/08/2020
							F19	Surgical Intervention	A070504	Premature Discharge of Battery	10/08/2020	10/08/2020
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	10/08/2020	10/08/2020
					CT321	Long-term surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	10/08/2020	10/08/2020
							F19	Surgical Intervention	A070504	Premature Discharge of Battery	10/08/2020	10/08/2020
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	10/08/2020	10/08/2020
					CT449	Generators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	10/08/2020	10/08/2020
							F19	Surgical Intervention	A070504	Premature Discharge of Battery	10/08/2020	10/08/2020
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	10/08/2020	10/08/2020
					CT981	Single use	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	10/08/2020	10/08/2020
							F19	Surgical Intervention	A070504	Premature Discharge of Battery	10/08/2020	10/08/2020
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	10/08/2020	10/08/2020
					CT983	Surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	10/08/2020	10/08/2020
							F19	Surgical Intervention	A070504	Premature Discharge of Battery	10/08/2020	10/08/2020
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	10/08/2020	10/08/2020
9608	2020	All other reportable incidents	37265	Dual-chamber implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	03/07/2020	03/07/2020
							F19	Surgical Intervention	A070504	Premature Discharge of Battery	03/07/2020	03/07/2020
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	03/07/2020	03/07/2020
9609	2020	All other reportable incidents	35852	Single-chamber implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	25/08/2020	25/08/2020
							F19	Surgical Intervention	A070504	Premature Discharge of Battery	25/08/2020	25/08/2020
9610	2020	All other reportable incidents	47265	Dual-chamber implantable pacemaker, rate-responsive	CT141	Electrodes/Leads and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	27/02/2020	27/02/2020
							F08	Hospitalization or Prolonged Hospitalization	A24	No Apparent Device Problem	27/02/2020	27/02/2020
							F12	Serious Injury/ Illness/ Impairment	A24	No Apparent Device Problem	27/02/2020	27/02/2020
							F1905	Device Revision or Replacement	A24	No Apparent Device Problem	27/02/2020	27/02/2020
							F24	Insufficient Health Impact Information	A24	No Apparent Device Problem	27/02/2020	27/02/2020
					CT145	Cardiovascular devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	27/02/2020	27/02/2020
							F08	Hospitalization or Prolonged Hospitalization	A24	No Apparent Device Problem	27/02/2020	27/02/2020
							F12	Serious Injury/ Illness/ Impairment	A24	No Apparent Device Problem	27/02/2020	27/02/2020
							F1905	Device Revision or Replacement	A24	No Apparent Device Problem	27/02/2020	27/02/2020
							F24	Insufficient Health Impact Information	A24	No Apparent Device Problem	27/02/2020	27/02/2020
					CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	27/02/2020	27/02/2020
							F08	Hospitalization or Prolonged Hospitalization	A24	No Apparent Device Problem	27/02/2020	27/02/2020
							F12	Serious Injury/ Illness/ Impairment	A24	No Apparent Device Problem	27/02/2020	27/02/2020
							F1905	Device Revision or Replacement	A24	No Apparent Device Problem	27/02/2020	27/02/2020
							F24	Insufficient Health Impact Information	A24	No Apparent Device Problem	27/02/2020	27/02/2020
					CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	27/02/2020	27/02/2020
							F08	Hospitalization or Prolonged Hospitalization	A24	No Apparent Device Problem	27/02/2020	27/02/2020
							F12	Serious Injury/ Illness/ Impairment	A24	No Apparent Device Problem	27/02/2020	27/02/2020
							F1905	Device Revision or Replacement	A24	No Apparent Device Problem	27/02/2020	27/02/2020
							F24	Insufficient Health Impact Information	A24	No Apparent Device Problem	27/02/2020	27/02/2020
					CT233	Surgical	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	27/02/2020	27/02/2020
							F08	Hospitalization or Prolonged Hospitalization	A24	No Apparent Device Problem	27/02/2020	27/02/2020
							F12	Serious Injury/ Illness/ Impairment	A24	No Apparent Device Problem	27/02/2020	27/02/2020
							F1905	Device Revision or Replacement	A24	No Apparent Device Problem	27/02/2020	27/02/2020
							F24	Insufficient Health Impact Information	A24	No Apparent Device Problem	27/02/2020	27/02/2020
					CT2406	Implantable	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	27/02/2020	27/02/2020
							F08	Hospitalization or Prolonged Hospitalization	A24	No Apparent Device Problem	27/02/2020	27/02/2020
							F12	Serious Injury/ Illness/ Impairment	A24	No Apparent Device Problem	27/02/2020	27/02/2020
							F1905	Device Revision or Replacement	A24	No Apparent Device Problem	27/02/2020	27/02/2020
							F24	Insufficient Health Impact Information	A24	No Apparent Device Problem	27/02/2020	27/02/2020
					CT2407	Active implantable	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	27/02/2020	27/02/2020
							F08	Hospitalization or Prolonged Hospitalization	A24	No Apparent Device Problem	27/02/2020	27/02/2020
							F12	Serious Injury/ Illness/ Impairment	A24	No Apparent Device Problem	27/02/2020	27/02/2020
							F1905	Device Revision or Replacement	A24	No Apparent Device Problem	27/02/2020	27/02/2020
							F24	Insufficient Health Impact Information	A24	No Apparent Device Problem	27/02/2020	27/02/2020
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	27/02/2020	27/02/2020
							F08	Hospitalization or Prolonged Hospitalization	A24	No Apparent Device Problem	27/02/2020	27/02/2020
							F12	Serious Injury/ Illness/ Impairment	A24	No Apparent Device Problem	27/02/2020	27/02/2020
							F1905	Device Revision or Replacement	A24	No Apparent Device Problem	27/02/2020	27/02/2020
							F24	Insufficient Health Impact Information	A24	No Apparent Device Problem	27/02/2020	27/02/2020
					CT2727	Active implant devices/ancillaries	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	27/02/2020	27/02/2020
							F08	Hospitalization or Prolonged Hospitalization	A24	No Apparent Device Problem	27/02/2020	27/02/2020
							F12	Serious Injury/ Illness/ Impairment	A24	No Apparent Device Problem	27/02/2020	27/02/2020
							F1905	Device Revision or Replacement	A24	No Apparent Device Problem	27/02/2020	27/02/2020
							F24	Insufficient Health Impact Information	A24	No Apparent Device Problem	27/02/2020	27/02/2020
					CT2750	Implantable cardiac electrical stimulators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	27/02/2020	27/02/2020
							F08	Hospitalization or Prolonged Hospitalization	A24	No Apparent Device Problem	27/02/2020	27/02/2020
							F12	Serious Injury/ Illness/ Impairment	A24	No Apparent Device Problem	27/02/2020	27/02/2020
							F1905	Device Revision or Replacement	A24	No Apparent Device Problem	27/02/2020	27/02/2020
							F24	Insufficient Health Impact Information	A24	No Apparent Device Problem	27/02/2020	27/02/2020
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	27/02/2020	27/02/2020
							F08	Hospitalization or Prolonged Hospitalization	A24	No Apparent Device Problem	27/02/2020	27/02/2020
							F12	Serious Injury/ Illness/ Impairment	A24	No Apparent Device Problem	27/02/2020	27/02/2020
							F1905	Device Revision or Replacement	A24	No Apparent Device Problem	27/02/2020	27/02/2020
							F24	Insufficient Health Impact Information	A24	No Apparent Device Problem	27/02/2020	27/02/2020
					CT2985	Active	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	27/02/2020	27/02/2020
							F08	Hospitalization or Prolonged Hospitalization	A24	No Apparent Device Problem	27/02/2020	27/02/2020
							F12	Serious Injury/ Illness/ Impairment	A24	No Apparent Device Problem	27/02/2020	27/02/2020
							F1905	Device Revision or Replacement	A24	No Apparent Device Problem	27/02/2020	27/02/2020
							F24	Insufficient Health Impact Information	A24	No Apparent Device Problem	27/02/2020	27/02/2020
					CT3001	Implantable metal	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	27/02/2020	27/02/2020
							F08	Hospitalization or Prolonged Hospitalization	A24	No Apparent Device Problem	27/02/2020	27/02/2020
							F12	Serious Injury/ Illness/ Impairment	A24	No Apparent Device Problem	27/02/2020	27/02/2020
							F1905	Device Revision or Replacement	A24	No Apparent Device Problem	27/02/2020	27/02/2020
							F24	Insufficient Health Impact Information	A24	No Apparent Device Problem	27/02/2020	27/02/2020

					CT3002	Implantable synthetic polymer	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	27/02/2020	27/02/2020
							F08	Hospitalization or Prolonged Hospitalization	A24	No Apparent Device Problem	27/02/2020	27/02/2020
							F12	Serious Injury/ Illness/ Impairment	A24	No Apparent Device Problem	27/02/2020	27/02/2020
							F1905	Device Revision or Replacement	A24	No Apparent Device Problem	27/02/2020	27/02/2020
							F24	Insufficient Health Impact Information	A24	No Apparent Device Problem	27/02/2020	27/02/2020
					CT3006	Implantable inorganic material	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	27/02/2020	27/02/2020
							F08	Hospitalization or Prolonged Hospitalization	A24	No Apparent Device Problem	27/02/2020	27/02/2020
							F12	Serious Injury/ Illness/ Impairment	A24	No Apparent Device Problem	27/02/2020	27/02/2020
							F1905	Device Revision or Replacement	A24	No Apparent Device Problem	27/02/2020	27/02/2020
							F24	Insufficient Health Impact Information	A24	No Apparent Device Problem	27/02/2020	27/02/2020
					CT314	Electrically powered	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	27/02/2020	27/02/2020
							F08	Hospitalization or Prolonged Hospitalization	A24	No Apparent Device Problem	27/02/2020	27/02/2020
							F12	Serious Injury/ Illness/ Impairment	A24	No Apparent Device Problem	27/02/2020	27/02/2020
							F1905	Device Revision or Replacement	A24	No Apparent Device Problem	27/02/2020	27/02/2020
							F24	Insufficient Health Impact Information	A24	No Apparent Device Problem	27/02/2020	27/02/2020
					CT321	Long-term surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	27/02/2020	27/02/2020
							F08	Hospitalization or Prolonged Hospitalization	A24	No Apparent Device Problem	27/02/2020	27/02/2020
							F12	Serious Injury/ Illness/ Impairment	A24	No Apparent Device Problem	27/02/2020	27/02/2020
							F1905	Device Revision or Replacement	A24	No Apparent Device Problem	27/02/2020	27/02/2020
							F24	Insufficient Health Impact Information	A24	No Apparent Device Problem	27/02/2020	27/02/2020
					CT449	Generators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	27/02/2020	27/02/2020
							F08	Hospitalization or Prolonged Hospitalization	A24	No Apparent Device Problem	27/02/2020	27/02/2020
							F12	Serious Injury/ Illness/ Impairment	A24	No Apparent Device Problem	27/02/2020	27/02/2020
							F1905	Device Revision or Replacement	A24	No Apparent Device Problem	27/02/2020	27/02/2020
							F24	Insufficient Health Impact Information	A24	No Apparent Device Problem	27/02/2020	27/02/2020
					CT981	Single use	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	27/02/2020	27/02/2020
							F08	Hospitalization or Prolonged Hospitalization	A24	No Apparent Device Problem	27/02/2020	27/02/2020
							F12	Serious Injury/ Illness/ Impairment	A24	No Apparent Device Problem	27/02/2020	27/02/2020
							F1905	Device Revision or Replacement	A24	No Apparent Device Problem	27/02/2020	27/02/2020
							F24	Insufficient Health Impact Information	A24	No Apparent Device Problem	27/02/2020	27/02/2020
					CT983	Surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	27/02/2020	27/02/2020
							F08	Hospitalization or Prolonged Hospitalization	A24	No Apparent Device Problem	27/02/2020	27/02/2020
							F12	Serious Injury/ Illness/ Impairment	A24	No Apparent Device Problem	27/02/2020	27/02/2020
							F1905	Device Revision or Replacement	A24	No Apparent Device Problem	27/02/2020	27/02/2020
							F24	Insufficient Health Impact Information	A24	No Apparent Device Problem	27/02/2020	27/02/2020
9611	2019	All other reportable incidents	60789	Intracardiac pacemaker	CT130	Catheters and associated devices	F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery		
					CT145	Cardiovascular devices	F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery		
					CT1684	Cardiac pacemakers and associated devices	F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery		
					CT1685	Cardiac pacemakers	F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery		
					CT233	Surgical	F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery		
					CT2406	Implantable	F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery		
					CT2407	Active implantable	F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery		
					CT257	Implantable cardiac pacemakers	F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery		
					CT2727	Active implant devices/ancillaries	F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery		
					CT2750	Implantable cardiac electrical stimulators and associated devices	F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery		
					CT2818	Catheterization/Cannulation/Fluid path devices	F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery		
					CT2866	Implantable cardiac electrical stimulators	F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery		
					CT2985	Active	F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery		
					CT3001	Implantable metal	F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery		
					CT3002	Implantable synthetic polymer	F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery		
					CT3006	Implantable inorganic material	F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery		
					CT314	Electrically powered	F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery		
					CT321	Long-term surgical invasive	F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery		
					CT981	Single use	F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery		
					CT983	Surgical invasive	F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery		
9612	2020	All other reportable incidents	47267	Single-chamber implantable pacemaker, rate-responsive	CT141	Electrodes/Leads and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	23/06/2020	23/06/2020
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	23/06/2020	23/06/2020
							F12	Serious Injury/ Illness/ Impairment	A070504	Premature Discharge of Battery	23/06/2020	23/06/2020
							F19	Surgical Intervention	A070504	Premature Discharge of Battery	23/06/2020	23/06/2020
					CT145	Cardiovascular devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	23/06/2020	23/06/2020
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	23/06/2020	23/06/2020
							F12	Serious Injury/ Illness/ Impairment	A070504	Premature Discharge of Battery	23/06/2020	23/06/2020
							F19	Surgical Intervention	A070504	Premature Discharge of Battery	23/06/2020	23/06/2020
					CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	23/06/2020	23/06/2020
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	23/06/2020	23/06/2020
							F12	Serious Injury/ Illness/ Impairment	A070504	Premature Discharge of Battery	23/06/2020	23/06/2020
							F19	Surgical Intervention	A070504	Premature Discharge of Battery	23/06/2020	23/06/2020
					CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	23/06/2020	23/06/2020
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	23/06/2020	23/06/2020
							F12	Serious Injury/ Illness/ Impairment	A070504	Premature Discharge of Battery	23/06/2020	23/06/2020
							F19	Surgical Intervention	A070504	Premature Discharge of Battery	23/06/2020	23/06/2020
					CT233	Surgical	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	23/06/2020	23/06/2020
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	23/06/2020	23/06/2020
							F12	Serious Injury/ Illness/ Impairment	A070504	Premature Discharge of Battery	23/06/2020	23/06/2020
							F19	Surgical Intervention	A070504	Premature Discharge of Battery	23/06/2020	23/06/2020
					CT2406	Implantable	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	23/06/2020	23/06/2020
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	23/06/2020	23/06/2020
							F12	Serious Injury/ Illness/ Impairment	A070504	Premature Discharge of Battery	23/06/2020	23/06/2020
							F19	Surgical Intervention	A070504	Premature Discharge of Battery	23/06/2020	23/06/2020
					CT2407	Active implantable	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	23/06/2020	23/06/2020
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	23/06/2020	23/06/2020
							F12	Serious Injury/ Illness/ Impairment	A070504	Premature Discharge of Battery	23/06/2020	23/06/2020
							F19	Surgical Intervention	A070504	Premature Discharge of Battery	23/06/2020	23/06/2020
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	23/06/2020	23/06/2020
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	23/06/2020	23/06/2020
							F12	Serious Injury/ Illness/ Impairment	A070504	Premature Discharge of Battery	23/06/2020	23/06/2020
							F19	Surgical Intervention	A070504	Premature Discharge of Battery	23/06/2020	23/06/2020

					CT2727	Active implant devices/ancillaries	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	23/06/2020	23/06/2020
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	23/06/2020	23/06/2020
							F12	Serious Injury/ Illness/ Impairment	A070504	Premature Discharge of Battery	23/06/2020	23/06/2020
							F19	Surgical Intervention	A070504	Premature Discharge of Battery	23/06/2020	23/06/2020
					CT2750	Implantable cardiac electrical stimulators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	23/06/2020	23/06/2020
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	23/06/2020	23/06/2020
							F12	Serious Injury/ Illness/ Impairment	A070504	Premature Discharge of Battery	23/06/2020	23/06/2020
							F19	Surgical Intervention	A070504	Premature Discharge of Battery	23/06/2020	23/06/2020
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	23/06/2020	23/06/2020
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	23/06/2020	23/06/2020
							F12	Serious Injury/ Illness/ Impairment	A070504	Premature Discharge of Battery	23/06/2020	23/06/2020
							F19	Surgical Intervention	A070504	Premature Discharge of Battery	23/06/2020	23/06/2020
					CT2985	Active	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	23/06/2020	23/06/2020
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	23/06/2020	23/06/2020
							F12	Serious Injury/ Illness/ Impairment	A070504	Premature Discharge of Battery	23/06/2020	23/06/2020
							F19	Surgical Intervention	A070504	Premature Discharge of Battery	23/06/2020	23/06/2020
					CT3001	Implantable metal	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	23/06/2020	23/06/2020
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	23/06/2020	23/06/2020
							F12	Serious Injury/ Illness/ Impairment	A070504	Premature Discharge of Battery	23/06/2020	23/06/2020
							F19	Surgical Intervention	A070504	Premature Discharge of Battery	23/06/2020	23/06/2020
					CT3002	Implantable synthetic polymer	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	23/06/2020	23/06/2020
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	23/06/2020	23/06/2020
							F12	Serious Injury/ Illness/ Impairment	A070504	Premature Discharge of Battery	23/06/2020	23/06/2020
							F19	Surgical Intervention	A070504	Premature Discharge of Battery	23/06/2020	23/06/2020
					CT3006	Implantable inorganic material	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	23/06/2020	23/06/2020
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	23/06/2020	23/06/2020
							F12	Serious Injury/ Illness/ Impairment	A070504	Premature Discharge of Battery	23/06/2020	23/06/2020
							F19	Surgical Intervention	A070504	Premature Discharge of Battery	23/06/2020	23/06/2020
					CT314	Electrically powered	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	23/06/2020	23/06/2020
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	23/06/2020	23/06/2020
							F12	Serious Injury/ Illness/ Impairment	A070504	Premature Discharge of Battery	23/06/2020	23/06/2020
							F19	Surgical Intervention	A070504	Premature Discharge of Battery	23/06/2020	23/06/2020
					CT321	Long-term surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	23/06/2020	23/06/2020
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	23/06/2020	23/06/2020
							F12	Serious Injury/ Illness/ Impairment	A070504	Premature Discharge of Battery	23/06/2020	23/06/2020
							F19	Surgical Intervention	A070504	Premature Discharge of Battery	23/06/2020	23/06/2020
					CT449	Generators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	23/06/2020	23/06/2020
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	23/06/2020	23/06/2020
							F12	Serious Injury/ Illness/ Impairment	A070504	Premature Discharge of Battery	23/06/2020	23/06/2020
							F19	Surgical Intervention	A070504	Premature Discharge of Battery	23/06/2020	23/06/2020
					CT981	Single use	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	23/06/2020	23/06/2020
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	23/06/2020	23/06/2020
							F12	Serious Injury/ Illness/ Impairment	A070504	Premature Discharge of Battery	23/06/2020	23/06/2020
							F19	Surgical Intervention	A070504	Premature Discharge of Battery	23/06/2020	23/06/2020
					CT983	Surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	23/06/2020	23/06/2020
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	23/06/2020	23/06/2020
							F12	Serious Injury/ Illness/ Impairment	A070504	Premature Discharge of Battery	23/06/2020	23/06/2020
							F19	Surgical Intervention	A070504	Premature Discharge of Battery	23/06/2020	23/06/2020
9613	2019	All other reportable incidents	47265	Dual-chamber implantable pacemaker, rate-responsive	CT141	Electrodes/Leads and associated devices	F14	Prolonged Episode of Care	A072201	High Impedance		
							F1905	Device Revision or Replacement	A072201	High Impedance		
					CT145	Cardiovascular devices	F14	Prolonged Episode of Care	A072201	High Impedance		
							F1905	Device Revision or Replacement	A072201	High Impedance		
					CT1684	Cardiac pacemakers and associated devices	F14	Prolonged Episode of Care	A072201	High Impedance		
							F1905	Device Revision or Replacement	A072201	High Impedance		
					CT1685	Cardiac pacemakers	F14	Prolonged Episode of Care	A072201	High Impedance		
							F1905	Device Revision or Replacement	A072201	High Impedance		
					CT233	Surgical	F14	Prolonged Episode of Care	A072201	High Impedance		
							F1905	Device Revision or Replacement	A072201	High Impedance		
					CT2406	Implantable	F14	Prolonged Episode of Care	A072201	High Impedance		
							F1905	Device Revision or Replacement	A072201	High Impedance		
					CT2407	Active implantable	F14	Prolonged Episode of Care	A072201	High Impedance		
							F1905	Device Revision or Replacement	A072201	High Impedance		
					CT257	Implantable cardiac pacemakers	F14	Prolonged Episode of Care	A072201	High Impedance		
							F1905	Device Revision or Replacement	A072201	High Impedance		
					CT2727	Active implant devices/ancillaries	F14	Prolonged Episode of Care	A072201	High Impedance		
							F1905	Device Revision or Replacement	A072201	High Impedance		
					CT2750	Implantable cardiac electrical stimulators and associated devices	F14	Prolonged Episode of Care	A072201	High Impedance		
							F1905	Device Revision or Replacement	A072201	High Impedance		
					CT2866	Implantable cardiac electrical stimulators	F14	Prolonged Episode of Care	A072201	High Impedance		
							F1905	Device Revision or Replacement	A072201	High Impedance		
					CT2985	Active	F14	Prolonged Episode of Care	A072201	High Impedance		
							F1905	Device Revision or Replacement	A072201	High Impedance		
					CT3001	Implantable metal	F14	Prolonged Episode of Care	A072201	High Impedance		
							F1905	Device Revision or Replacement	A072201	High Impedance		
					CT3002	Implantable synthetic polymer	F14	Prolonged Episode of Care	A072201	High Impedance		
							F1905	Device Revision or Replacement	A072201	High Impedance		
					CT3006	Implantable inorganic material	F14	Prolonged Episode of Care	A072201	High Impedance		
							F1905	Device Revision or Replacement	A072201	High Impedance		
					CT314	Electrically powered	F14	Prolonged Episode of Care	A072201	High Impedance		
							F1905	Device Revision or Replacement	A072201	High Impedance		
					CT321	Long-term surgical invasive	F14	Prolonged Episode of Care	A072201	High Impedance		
							F1905	Device Revision or Replacement	A072201	High Impedance		
					CT449	Generators and associated devices	F14	Prolonged Episode of Care	A072201	High Impedance		
							F1905	Device Revision or Replacement	A072201	High Impedance		
					CT981	Single use	F14	Prolonged Episode of Care	A072201	High Impedance		
							F1905	Device Revision or Replacement	A072201	High Impedance		
					CT983	Surgical invasive	F14	Prolonged Episode of Care	A072201	High Impedance		
							F1905	Device Revision or Replacement	A072201	High Impedance		

9614	2019	All other reportable incidents	35852	Single-chamber implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	06/09/2019		
9615	2019	All other reportable incidents	36241	Pacemaker/defibrillator lead	CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A072201	High Impedance			
9616	2019	All other reportable incidents	35852	Single-chamber implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A	Annex A			
9617	2019	All other reportable incidents	36241	Pacemaker/defibrillator lead	CT1684	Cardiac pacemakers and associated devices	F1905	Device Revision or Replacement	A070909	Oversensing			
9618	2019	All other reportable incidents	35852	Single-chamber implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A072201	High Impedance			
9619	2019	All other reportable incidents	35852	Single-chamber implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A072201	High Impedance			
9620	2019	All other reportable incidents	35852	Single-chamber implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A071302	Inappropriate/Inadequate Shock/Stimulation			
9621	2020	All other reportable incidents	47267	Single-chamber implantable pacemaker, rate-responsive	CT141	Electrodes/Leads and associated devices	E0112	Dizziness	A070103	Intermittent Capture			
							E011903	Syncope/Fainting	A070103	Intermittent Capture			
							F23	Unexpected Medical Intervention	A070103	Intermittent Capture			
					CT145	Cardiovascular devices	E0112	Dizziness	A070103	Intermittent Capture			
							E011903	Syncope/Fainting	A070103	Intermittent Capture			
							F23	Unexpected Medical Intervention	A070103	Intermittent Capture			
					CT1684	Cardiac pacemakers and associated devices	E0112	Dizziness	A070103	Intermittent Capture			
							E011903	Syncope/Fainting	A070103	Intermittent Capture			
							F23	Unexpected Medical Intervention	A070103	Intermittent Capture			
					CT1685	Cardiac pacemakers	E0112	Dizziness	A070103	Intermittent Capture			
							E011903	Syncope/Fainting	A070103	Intermittent Capture			
							F23	Unexpected Medical Intervention	A070103	Intermittent Capture			
					CT233	Surgical	E0112	Dizziness	A070103	Intermittent Capture			
							E011903	Syncope/Fainting	A070103	Intermittent Capture			
							F23	Unexpected Medical Intervention	A070103	Intermittent Capture			
					CT2406	Implantable	E0112	Dizziness	A070103	Intermittent Capture			
							E011903	Syncope/Fainting	A070103	Intermittent Capture			
							F23	Unexpected Medical Intervention	A070103	Intermittent Capture			
					CT2407	Active implantable	E0112	Dizziness	A070103	Intermittent Capture			
							E011903	Syncope/Fainting	A070103	Intermittent Capture			
							F23	Unexpected Medical Intervention	A070103	Intermittent Capture			
					CT257	Implantable cardiac pacemakers	E0112	Dizziness	A070103	Intermittent Capture			
							E011903	Syncope/Fainting	A070103	Intermittent Capture			
							F23	Unexpected Medical Intervention	A070103	Intermittent Capture			
					CT2727	Active implant devices/ancillaries	E0112	Dizziness	A070103	Intermittent Capture			
							E011903	Syncope/Fainting	A070103	Intermittent Capture			
							F23	Unexpected Medical Intervention	A070103	Intermittent Capture			
					CT2750	Implantable cardiac electrical stimulators and associated devices	E0112	Dizziness	A070103	Intermittent Capture			
							E011903	Syncope/Fainting	A070103	Intermittent Capture			
							F23	Unexpected Medical Intervention	A070103	Intermittent Capture			
					CT2866	Implantable cardiac electrical stimulators	E0112	Dizziness	A070103	Intermittent Capture			
							E011903	Syncope/Fainting	A070103	Intermittent Capture			
							F23	Unexpected Medical Intervention	A070103	Intermittent Capture			
					CT2985	Active	E0112	Dizziness	A070103	Intermittent Capture			
							E011903	Syncope/Fainting	A070103	Intermittent Capture			
							F23	Unexpected Medical Intervention	A070103	Intermittent Capture			
					CT3001	Implantable metal	E0112	Dizziness	A070103	Intermittent Capture			
							E011903	Syncope/Fainting	A070103	Intermittent Capture			
							F23	Unexpected Medical Intervention	A070103	Intermittent Capture			
					CT3002	Implantable synthetic polymer	E0112	Dizziness	A070103	Intermittent Capture			
							E011903	Syncope/Fainting	A070103	Intermittent Capture			
							F23	Unexpected Medical Intervention	A070103	Intermittent Capture			
					CT3006	Implantable inorganic material	E0112	Dizziness	A070103	Intermittent Capture			
							E011903	Syncope/Fainting	A070103	Intermittent Capture			
							F23	Unexpected Medical Intervention	A070103	Intermittent Capture			
					CT314	Electrically powered	E0112	Dizziness	A070103	Intermittent Capture			
							E011903	Syncope/Fainting	A070103	Intermittent Capture			
							F23	Unexpected Medical Intervention	A070103	Intermittent Capture			
					CT321	Long-term surgical invasive	E0112	Dizziness	A070103	Intermittent Capture			
							E011903	Syncope/Fainting	A070103	Intermittent Capture			
							F23	Unexpected Medical Intervention	A070103	Intermittent Capture			
					CT449	Generators and associated devices	E0112	Dizziness	A070103	Intermittent Capture			
							E011903	Syncope/Fainting	A070103	Intermittent Capture			
							F23	Unexpected Medical Intervention	A070103	Intermittent Capture			
					CT981	Single use	E0112	Dizziness	A070103	Intermittent Capture			
							E011903	Syncope/Fainting	A070103	Intermittent Capture			
							F23	Unexpected Medical Intervention	A070103	Intermittent Capture			
					CT983	Surgical invasive	E0112	Dizziness	A070103	Intermittent Capture			
							E011903	Syncope/Fainting	A070103	Intermittent Capture			
							F23	Unexpected Medical Intervention	A070103	Intermittent Capture			
9622	2020	All other reportable incidents	35852	Single-chamber implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A070909	Oversensing			
									A090801	Signal Artifact/Noise			
									A2304	Off-Label Use			
							F26	No Health Consequences or Impact	A070909	Oversensing			
									A090801	Signal Artifact/Noise			
									A2304	Off-Label Use			
9623	2020	All other reportable incidents	47265	Dual-chamber implantable pacemaker, rate-responsive	CT141	Electrodes/Leads and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070909	Oversensing			
							F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A070909	Oversensing			
							F15	Recognized Device or Procedural Complication	A070909	Oversensing			
							F26	No Health Consequences or Impact	A070909	Oversensing			
					CT145	Cardiovascular devices	E2403	No Clinical Signs, Symptoms or Conditions	A070909	Oversensing			
							F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A070909	Oversensing			
							F15	Recognized Device or Procedural Complication	A070909	Oversensing			
							F26	No Health Consequences or Impact	A070909	Oversensing			
					CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070909	Oversensing			
							F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A070909	Oversensing			
							F15	Recognized Device or Procedural Complication	A070909	Oversensing			
							F26	No Health Consequences or Impact	A070909	Oversensing			
					CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A070909	Oversensing			
							F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A070909	Oversensing			
							F15	Recognized Device or Procedural Complication	A070909	Oversensing			

							F26	No Health Consequences or Impact	A070909	Oversensing		
					CT233	Surgical	E2403	No Clinical Signs, Symptoms or Conditions	A070909	Oversensing		
							F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A070909	Oversensing		
							F15	Recognized Device or Procedural Complication	A070909	Oversensing		
							F26	No Health Consequences or Impact	A070909	Oversensing		
					CT2406	Implantable	E2403	No Clinical Signs, Symptoms or Conditions	A070909	Oversensing		
							F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A070909	Oversensing		
							F15	Recognized Device or Procedural Complication	A070909	Oversensing		
							F26	No Health Consequences or Impact	A070909	Oversensing		
					CT2407	Active implantable	E2403	No Clinical Signs, Symptoms or Conditions	A070909	Oversensing		
							F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A070909	Oversensing		
							F15	Recognized Device or Procedural Complication	A070909	Oversensing		
							F26	No Health Consequences or Impact	A070909	Oversensing		
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A070909	Oversensing		
							F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A070909	Oversensing		
							F15	Recognized Device or Procedural Complication	A070909	Oversensing		
							F26	No Health Consequences or Impact	A070909	Oversensing		
					CT2727	Active implant devices/ancillaries	E2403	No Clinical Signs, Symptoms or Conditions	A070909	Oversensing		
							F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A070909	Oversensing		
							F15	Recognized Device or Procedural Complication	A070909	Oversensing		
							F26	No Health Consequences or Impact	A070909	Oversensing		
					CT2750	Implantable cardiac electrical stimulators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070909	Oversensing		
							F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A070909	Oversensing		
							F15	Recognized Device or Procedural Complication	A070909	Oversensing		
							F26	No Health Consequences or Impact	A070909	Oversensing		
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A070909	Oversensing		
							F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A070909	Oversensing		
							F15	Recognized Device or Procedural Complication	A070909	Oversensing		
							F26	No Health Consequences or Impact	A070909	Oversensing		
					CT2985	Active	E2403	No Clinical Signs, Symptoms or Conditions	A070909	Oversensing		
							F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A070909	Oversensing		
							F15	Recognized Device or Procedural Complication	A070909	Oversensing		
							F26	No Health Consequences or Impact	A070909	Oversensing		
					CT3001	Implantable metal	E2403	No Clinical Signs, Symptoms or Conditions	A070909	Oversensing		
							F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A070909	Oversensing		
							F15	Recognized Device or Procedural Complication	A070909	Oversensing		
							F26	No Health Consequences or Impact	A070909	Oversensing		
					CT3002	Implantable synthetic polymer	E2403	No Clinical Signs, Symptoms or Conditions	A070909	Oversensing		
							F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A070909	Oversensing		
							F15	Recognized Device or Procedural Complication	A070909	Oversensing		
							F26	No Health Consequences or Impact	A070909	Oversensing		
					CT3006	Implantable inorganic material	E2403	No Clinical Signs, Symptoms or Conditions	A070909	Oversensing		
							F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A070909	Oversensing		
							F15	Recognized Device or Procedural Complication	A070909	Oversensing		
							F26	No Health Consequences or Impact	A070909	Oversensing		
					CT314	Electrically powered	E2403	No Clinical Signs, Symptoms or Conditions	A070909	Oversensing		
							F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A070909	Oversensing		
							F15	Recognized Device or Procedural Complication	A070909	Oversensing		
							F26	No Health Consequences or Impact	A070909	Oversensing		
					CT321	Long-term surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A070909	Oversensing		
							F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A070909	Oversensing		
							F15	Recognized Device or Procedural Complication	A070909	Oversensing		
							F26	No Health Consequences or Impact	A070909	Oversensing		
					CT449	Generators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070909	Oversensing		
							F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A070909	Oversensing		
							F15	Recognized Device or Procedural Complication	A070909	Oversensing		
							F26	No Health Consequences or Impact	A070909	Oversensing		
					CT981	Single use	E2403	No Clinical Signs, Symptoms or Conditions	A070909	Oversensing		
							F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A070909	Oversensing		
							F15	Recognized Device or Procedural Complication	A070909	Oversensing		
							F26	No Health Consequences or Impact	A070909	Oversensing		
					CT983	Surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A070909	Oversensing		
							F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A070909	Oversensing		
							F15	Recognized Device or Procedural Complication	A070909	Oversensing		
							F26	No Health Consequences or Impact	A070909	Oversensing		
9624	2019	All other reportable incidents	35852	Single-chamber implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E0503	Embolism/Embolus	A24	No Apparent Device Problem		
9625	2020	All other reportable incidents	35852	Single-chamber implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	25/08/2020	25/08/2020
							F19	Surgical Intervention	A070504	Premature Discharge of Battery	25/08/2020	25/08/2020
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	25/08/2020	25/08/2020
9626	2019	All other reportable incidents	35852	Single-chamber implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A071302	Inappropriate/Inadequate Shock/Stimulation		
9627	2019	All other reportable incidents	35852	Single-chamber implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A070909	Oversensing		
9628	2020	All other reportable incidents	35852	Single-chamber implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	19/07/2020	19/07/2020
							F1901	Additional Surgery	A070504	Premature Discharge of Battery	19/07/2020	19/07/2020
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	19/07/2020	19/07/2020
9629	2020	All other reportable incidents	35223	Endocardial/interventricular septal pacing lead	CT1684	Cardiac pacemakers and associated devices	E011903	Syncope/Fainting	A0401	Break	06/07/2020	06/07/2020
									A070901	Decreased Sensitivity	06/07/2020	06/07/2020
									A0722	Impedance Problem	06/07/2020	06/07/2020
							E060109	Tachycardia	A0401	Break	06/07/2020	06/07/2020
									A070901	Decreased Sensitivity	06/07/2020	06/07/2020
									A0722	Impedance Problem	06/07/2020	06/07/2020
							E2007	Fall	A0401	Break	06/07/2020	06/07/2020
									A070901	Decreased Sensitivity	06/07/2020	06/07/2020
									A0722	Impedance Problem	06/07/2020	06/07/2020
							E2401	Insufficient Information	A0401	Break	06/07/2020	06/07/2020
									A070901	Decreased Sensitivity	06/07/2020	06/07/2020
									A0722	Impedance Problem	06/07/2020	06/07/2020
							F02	Death	A0401	Break	06/07/2020	06/07/2020
									A070901	Decreased Sensitivity	06/07/2020	06/07/2020

									A0722	Impedance Problem	06/07/2020	06/07/2020
9630	2019	All other reportable incidents	47205	Cardiac pulse generator programmer	CT1684	Cardiac pacemakers and associated devices	E0112	Dizziness	A1303	Intermittent Communication Failure		
9631	2019	All other reportable incidents	36241	Pacemaker/defibrillator lead	CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A0709	Device Sensing Problem		
9632	2019	Death	17882	Physiologic-monitoring defibrillation system	CT1684	Cardiac pacemakers and associated devices	F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A1302	Failure to Transmit Record		
9633	2019	All other reportable incidents	35223	Endocardial/interventricular septal pacing lead	CT1684	Cardiac pacemakers and associated devices	F1905	Device Revision or Replacement	A072201	High Impedance		
9634	2020	All other reportable incidents	35223	Endocardial/interventricular septal pacing lead	CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A040101	Fracture		
									A070909	Oversensing		
							F1901	Additional Surgery	A040101	Fracture		
									A070909	Oversensing		
							F1905	Device Revision or Replacement	A040101	Fracture		
									A070909	Oversensing		
9635	2020	All other reportable incidents	47270	Cardiac resynchronization therapy implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A071302	Inappropriate/Inadequate Shock/Stimulation		
							F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A071302	Inappropriate/Inadequate Shock/Stimulation		
							F26	No Health Consequences or Impact	A071302	Inappropriate/Inadequate Shock/Stimulation		
9636	2020	All other reportable incidents	47270	Cardiac resynchronization therapy implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	21/07/2020	21/07/2020
									A071102	Failure to Interrogate	21/07/2020	21/07/2020
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	21/07/2020	21/07/2020
									A071102	Failure to Interrogate	21/07/2020	21/07/2020
9637	2020	All other reportable incidents	47265	Dual-chamber implantable pacemaker, rate-responsive	CT141	Electrodes/Leads and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT145	Cardiovascular devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT233	Surgical	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT2406	Implantable	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT2407	Active implantable	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT2727	Active implant devices/ancillaries	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT2750	Implantable cardiac electrical stimulators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT2985	Active	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT3001	Implantable metal	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT3002	Implantable synthetic polymer	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT3006	Implantable inorganic material	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT314	Electrically powered	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT321	Long-term surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT449	Generators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT981	Single use	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT983	Surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
9638	2020	All other reportable incidents	36241	Pacemaker/defibrillator lead	CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A05	Mechanical Problem		
							F26	No Health Consequences or Impact	A05	Mechanical Problem		
9639	2020	All other reportable incidents	47265	Dual-chamber implantable pacemaker, rate-responsive	CT141	Electrodes/Leads and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT145	Cardiovascular devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT233	Surgical	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT2406	Implantable	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT2407	Active implantable	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT2727	Active implant devices/ancillaries	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT2750	Implantable cardiac electrical stimulators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT2985	Active	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT3001	Implantable metal	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT3002	Implantable synthetic polymer	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		

							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT3006	Implantable inorganic material	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT314	Electrically powered	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT321	Long-term surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT449	Generators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT981	Single use	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT983	Surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
9640	2020	All other reportable incidents	36241	Pacemaker/defibrillator lead	CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A072201	High Impedance		
							F26	No Health Consequences or Impact	A072201	High Impedance		
9641	2020	All other reportable incidents	35852	Single-chamber implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E2104	Electric Shock	A070909	Oversensing		
									A071302	Inappropriate/Inadequate Shock/Stimulation		
							E2403	No Clinical Signs, Symptoms or Conditions	A070909	Oversensing		
									A071302	Inappropriate/Inadequate Shock/Stimulation		
							F26	No Health Consequences or Impact	A070909	Oversensing		
									A071302	Inappropriate/Inadequate Shock/Stimulation		
9642	2020	All other reportable incidents	36241	Pacemaker/defibrillator lead	CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A072201	High Impedance		
							F26	No Health Consequences or Impact	A072201	High Impedance		
9643	2019	All other reportable incidents	36241	Pacemaker/defibrillator lead	CT1684	Cardiac pacemakers and associated devices	F1905	Device Revision or Replacement	A072201	High Impedance	13/08/2019	
9644	2019	All other reportable incidents	36241	Pacemaker/defibrillator lead	CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070909	Oversensing		
9645	2019	All other reportable incidents	35852	Single-chamber implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A070909	Oversensing	21/01/2020	21/01/2020
9646	2019	All other reportable incidents	47265	Dual-chamber implantable pacemaker, rate-responsive	CT141	Electrodes/Leads and associated devices	F1905	Device Revision or Replacement	A071205	Pacemaker Found in Back-Up Mode	14/08/2019	
					CT145	Cardiovascular devices	F1905	Device Revision or Replacement	A071205	Pacemaker Found in Back-Up Mode	14/08/2019	
					CT1684	Cardiac pacemakers and associated devices	F1905	Device Revision or Replacement	A071205	Pacemaker Found in Back-Up Mode	14/08/2019	
					CT1685	Cardiac pacemakers	F1905	Device Revision or Replacement	A071205	Pacemaker Found in Back-Up Mode	14/08/2019	
					CT233	Surgical	F1905	Device Revision or Replacement	A071205	Pacemaker Found in Back-Up Mode	14/08/2019	
					CT2406	Implantable	F1905	Device Revision or Replacement	A071205	Pacemaker Found in Back-Up Mode	14/08/2019	
					CT2407	Active implantable	F1905	Device Revision or Replacement	A071205	Pacemaker Found in Back-Up Mode	14/08/2019	
					CT257	Implantable cardiac pacemakers	F1905	Device Revision or Replacement	A071205	Pacemaker Found in Back-Up Mode	14/08/2019	
					CT2727	Active implant devices/ancillaries	F1905	Device Revision or Replacement	A071205	Pacemaker Found in Back-Up Mode	14/08/2019	
					CT2750	Implantable cardiac electrical stimulators and associated devices	F1905	Device Revision or Replacement	A071205	Pacemaker Found in Back-Up Mode	14/08/2019	
					CT2866	Implantable cardiac electrical stimulators	F1905	Device Revision or Replacement	A071205	Pacemaker Found in Back-Up Mode	14/08/2019	
					CT2985	Active	F1905	Device Revision or Replacement	A071205	Pacemaker Found in Back-Up Mode	14/08/2019	
					CT3001	Implantable metal	F1905	Device Revision or Replacement	A071205	Pacemaker Found in Back-Up Mode	14/08/2019	
					CT3002	Implantable synthetic polymer	F1905	Device Revision or Replacement	A071205	Pacemaker Found in Back-Up Mode	14/08/2019	
					CT3006	Implantable inorganic material	F1905	Device Revision or Replacement	A071205	Pacemaker Found in Back-Up Mode	14/08/2019	
					CT314	Electrically powered	F1905	Device Revision or Replacement	A071205	Pacemaker Found in Back-Up Mode	14/08/2019	
					CT321	Long-term surgical invasive	F1905	Device Revision or Replacement	A071205	Pacemaker Found in Back-Up Mode	14/08/2019	
					CT449	Generators and associated devices	F1905	Device Revision or Replacement	A071205	Pacemaker Found in Back-Up Mode	14/08/2019	
					CT981	Single use	F1905	Device Revision or Replacement	A071205	Pacemaker Found in Back-Up Mode	14/08/2019	
					CT983	Surgical invasive	F1905	Device Revision or Replacement	A071205	Pacemaker Found in Back-Up Mode	14/08/2019	
9647	2019	All other reportable incidents	35852	Single-chamber implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A070909	Oversensing		
9648	2019	All other reportable incidents	36241	Pacemaker/defibrillator lead	CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A0709	Device Sensing Problem	12/08/2019	12/08/2019
9649	2019	All other reportable incidents	17882	Physiologic-monitoring defibrillation system	CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070803	Failure to Power Up		
							F27	Problem Identified During Non-Clinical Procedure	A070803	Failure to Power Up		
9650	2019	All other reportable incidents	17882	Physiologic-monitoring defibrillation system	CT1684	Cardiac pacemakers and associated devices	F05	Delay to Treatment/ Therapy	A11	Computer Software Problem		
9651	2019	All other reportable incidents	35852	Single-chamber implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
9652	2020	All other reportable incidents	36241	Pacemaker/defibrillator lead	CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A05	Mechanical Problem	24/07/2020	24/07/2020
									A051201	Device Dislodged or Dislocated	24/07/2020	24/07/2020
									A070102	High Capture Threshold	24/07/2020	24/07/2020
							F12	Serious Injury/ Illness/ Impairment	A05	Mechanical Problem	24/07/2020	24/07/2020
									A051201	Device Dislodged or Dislocated	24/07/2020	24/07/2020
									A070102	High Capture Threshold	24/07/2020	24/07/2020
							F1905	Device Revision or Replacement	A05	Mechanical Problem	24/07/2020	24/07/2020
									A051201	Device Dislodged or Dislocated	24/07/2020	24/07/2020
									A070102	High Capture Threshold	24/07/2020	24/07/2020
9653	2020	All other reportable incidents	36241	Pacemaker/defibrillator lead	CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A051201	Device Dislodged or Dislocated		
									A070102	High Capture Threshold		
							F12	Serious Injury/ Illness/ Impairment	A051201	Device Dislodged or Dislocated		
									A070102	High Capture Threshold		
							F1904	Device Repositioning	A051201	Device Dislodged or Dislocated		
									A070102	High Capture Threshold		
9654	2019	All other reportable incidents	47265	Dual-chamber implantable pacemaker, rate-responsive	CT141	Electrodes/Leads and associated devices	F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	10/01/2020	10/01/2020
					CT145	Cardiovascular devices	F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	10/01/2020	10/01/2020
					CT1684	Cardiac pacemakers and associated devices	F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	10/01/2020	10/01/2020
					CT1685	Cardiac pacemakers	F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	10/01/2020	10/01/2020
					CT233	Surgical	F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	10/01/2020	10/01/2020
					CT2406	Implantable	F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	10/01/2020	10/01/2020
					CT2407	Active implantable	F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	10/01/2020	10/01/2020
					CT257	Implantable cardiac pacemakers	F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	10/01/2020	10/01/2020
					CT2727	Active implant devices/ancillaries	F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	10/01/2020	10/01/2020
					CT2750	Implantable cardiac electrical stimulators and associated devices	F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	10/01/2020	10/01/2020
					CT2866	Implantable cardiac electrical stimulators	F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	10/01/2020	10/01/2020
					CT2985	Active	F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	10/01/2020	10/01/2020
					CT3001	Implantable metal	F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	10/01/2020	10/01/2020
					CT3002	Implantable synthetic polymer	F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	10/01/2020	10/01/2020
					CT3006	Implantable inorganic material	F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	10/01/2020	10/01/2020
					CT314	Electrically powered	F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	10/01/2020	10/01/2020
					CT321	Long-term surgical invasive	F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	10/01/2020	10/01/2020
					CT449	Generators and associated devices	F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	10/01/2020	10/01/2020
					CT981	Single use	F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	10/01/2020	10/01/2020

					CT983	Surgical invasive	F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	10/01/2020	10/01/2020
9655	2019	All other reportable incidents	47265	Dual-chamber implantable pacemaker, rate-responsive	CT141	Electrodes/Leads and associated devices	F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	28/08/2019	28/08/2019
					CT145	Cardiovascular devices	F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	28/08/2019	28/08/2019
					CT1684	Cardiac pacemakers and associated devices	F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	28/08/2019	28/08/2019
					CT1685	Cardiac pacemakers	F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	28/08/2019	28/08/2019
					CT233	Surgical	F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	28/08/2019	28/08/2019
					CT2406	Implantable	F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	28/08/2019	28/08/2019
					CT2407	Active implantable	F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	28/08/2019	28/08/2019
					CT257	Implantable cardiac pacemakers	F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	28/08/2019	28/08/2019
					CT2727	Active implant devices/ancillaries	F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	28/08/2019	28/08/2019
					CT2750	Implantable cardiac electrical stimulators and associated devices	F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	28/08/2019	28/08/2019
					CT2866	Implantable cardiac electrical stimulators	F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	28/08/2019	28/08/2019
					CT2985	Active	F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	28/08/2019	28/08/2019
					CT3001	Implantable metal	F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	28/08/2019	28/08/2019
					CT3002	Implantable synthetic polymer	F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	28/08/2019	28/08/2019
					CT3006	Implantable inorganic material	F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	28/08/2019	28/08/2019
					CT314	Electrically powered	F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	28/08/2019	28/08/2019
					CT321	Long-term surgical invasive	F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	28/08/2019	28/08/2019
					CT449	Generators and associated devices	F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	28/08/2019	28/08/2019
					CT981	Single use	F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	28/08/2019	28/08/2019
					CT983	Surgical invasive	F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	28/08/2019	28/08/2019
9656	2020	All other reportable incidents	47265	Dual-chamber implantable pacemaker, rate-responsive	CT141	Electrodes/Leads and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070909	Oversensing		
							F15	Recognized Device or Procedural Complication	A070909	Oversensing		
							F26	No Health Consequences or Impact	A070909	Oversensing		
					CT145	Cardiovascular devices	E2403	No Clinical Signs, Symptoms or Conditions	A070909	Oversensing		
							F15	Recognized Device or Procedural Complication	A070909	Oversensing		
							F26	No Health Consequences or Impact	A070909	Oversensing		
					CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070909	Oversensing		
							F15	Recognized Device or Procedural Complication	A070909	Oversensing		
							F26	No Health Consequences or Impact	A070909	Oversensing		
					CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A070909	Oversensing		
							F15	Recognized Device or Procedural Complication	A070909	Oversensing		
							F26	No Health Consequences or Impact	A070909	Oversensing		
					CT233	Surgical	E2403	No Clinical Signs, Symptoms or Conditions	A070909	Oversensing		
							F15	Recognized Device or Procedural Complication	A070909	Oversensing		
							F26	No Health Consequences or Impact	A070909	Oversensing		
					CT2406	Implantable	E2403	No Clinical Signs, Symptoms or Conditions	A070909	Oversensing		
							F15	Recognized Device or Procedural Complication	A070909	Oversensing		
							F26	No Health Consequences or Impact	A070909	Oversensing		
					CT2407	Active implantable	E2403	No Clinical Signs, Symptoms or Conditions	A070909	Oversensing		
							F15	Recognized Device or Procedural Complication	A070909	Oversensing		
							F26	No Health Consequences or Impact	A070909	Oversensing		
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A070909	Oversensing		
							F15	Recognized Device or Procedural Complication	A070909	Oversensing		
							F26	No Health Consequences or Impact	A070909	Oversensing		
					CT2727	Active implant devices/ancillaries	E2403	No Clinical Signs, Symptoms or Conditions	A070909	Oversensing		
							F15	Recognized Device or Procedural Complication	A070909	Oversensing		
							F26	No Health Consequences or Impact	A070909	Oversensing		
					CT2750	Implantable cardiac electrical stimulators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070909	Oversensing		
							F15	Recognized Device or Procedural Complication	A070909	Oversensing		
							F26	No Health Consequences or Impact	A070909	Oversensing		
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A070909	Oversensing		
							F15	Recognized Device or Procedural Complication	A070909	Oversensing		
							F26	No Health Consequences or Impact	A070909	Oversensing		
					CT2985	Active	E2403	No Clinical Signs, Symptoms or Conditions	A070909	Oversensing		
							F15	Recognized Device or Procedural Complication	A070909	Oversensing		
							F26	No Health Consequences or Impact	A070909	Oversensing		
					CT3001	Implantable metal	E2403	No Clinical Signs, Symptoms or Conditions	A070909	Oversensing		
							F15	Recognized Device or Procedural Complication	A070909	Oversensing		
							F26	No Health Consequences or Impact	A070909	Oversensing		
					CT3002	Implantable synthetic polymer	E2403	No Clinical Signs, Symptoms or Conditions	A070909	Oversensing		
							F15	Recognized Device or Procedural Complication	A070909	Oversensing		
							F26	No Health Consequences or Impact	A070909	Oversensing		
					CT3006	Implantable inorganic material	E2403	No Clinical Signs, Symptoms or Conditions	A070909	Oversensing		
							F15	Recognized Device or Procedural Complication	A070909	Oversensing		
							F26	No Health Consequences or Impact	A070909	Oversensing		
					CT314	Electrically powered	E2403	No Clinical Signs, Symptoms or Conditions	A070909	Oversensing		
							F15	Recognized Device or Procedural Complication	A070909	Oversensing		
							F26	No Health Consequences or Impact	A070909	Oversensing		
					CT321	Long-term surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A070909	Oversensing		
							F15	Recognized Device or Procedural Complication	A070909	Oversensing		
							F26	No Health Consequences or Impact	A070909	Oversensing		
					CT449	Generators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070909	Oversensing		
							F15	Recognized Device or Procedural Complication	A070909	Oversensing		
							F26	No Health Consequences or Impact	A070909	Oversensing		
					CT981	Single use	E2403	No Clinical Signs, Symptoms or Conditions	A070909	Oversensing		
							F15	Recognized Device or Procedural Complication	A070909	Oversensing		
							F26	No Health Consequences or Impact	A070909	Oversensing		
					CT983	Surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A070909	Oversensing		
							F15	Recognized Device or Procedural Complication	A070909	Oversensing		
							F26	No Health Consequences or Impact	A070909	Oversensing		
9657	2020	All other reportable incidents	47267	Single-chamber implantable pacemaker, rate-responsive	CT141	Electrodes/Leads and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070909	Oversensing		
							F15	Recognized Device or Procedural Complication	A070909	Oversensing		
							F26	No Health Consequences or Impact	A070909	Oversensing		
					CT145	Cardiovascular devices	E2403	No Clinical Signs, Symptoms or Conditions	A070909	Oversensing		
							F15	Recognized Device or Procedural Complication	A070909	Oversensing		
							F26	No Health Consequences or Impact	A070909	Oversensing		

					CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070909	Oversensing		
							F15	Recognized Device or Procedural Complication	A070909	Oversensing		
							F26	No Health Consequences or Impact	A070909	Oversensing		
					CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A070909	Oversensing		
							F15	Recognized Device or Procedural Complication	A070909	Oversensing		
							F26	No Health Consequences or Impact	A070909	Oversensing		
					CT233	Surgical	E2403	No Clinical Signs, Symptoms or Conditions	A070909	Oversensing		
							F15	Recognized Device or Procedural Complication	A070909	Oversensing		
							F26	No Health Consequences or Impact	A070909	Oversensing		
					CT2406	Implantable	E2403	No Clinical Signs, Symptoms or Conditions	A070909	Oversensing		
							F15	Recognized Device or Procedural Complication	A070909	Oversensing		
							F26	No Health Consequences or Impact	A070909	Oversensing		
					CT2407	Active implantable	E2403	No Clinical Signs, Symptoms or Conditions	A070909	Oversensing		
							F15	Recognized Device or Procedural Complication	A070909	Oversensing		
							F26	No Health Consequences or Impact	A070909	Oversensing		
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A070909	Oversensing		
							F15	Recognized Device or Procedural Complication	A070909	Oversensing		
							F26	No Health Consequences or Impact	A070909	Oversensing		
					CT2727	Active implant devices/ancillaries	E2403	No Clinical Signs, Symptoms or Conditions	A070909	Oversensing		
							F15	Recognized Device or Procedural Complication	A070909	Oversensing		
							F26	No Health Consequences or Impact	A070909	Oversensing		
					CT2750	Implantable cardiac electrical stimulators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070909	Oversensing		
							F15	Recognized Device or Procedural Complication	A070909	Oversensing		
							F26	No Health Consequences or Impact	A070909	Oversensing		
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A070909	Oversensing		
							F15	Recognized Device or Procedural Complication	A070909	Oversensing		
							F26	No Health Consequences or Impact	A070909	Oversensing		
					CT2985	Active	E2403	No Clinical Signs, Symptoms or Conditions	A070909	Oversensing		
							F15	Recognized Device or Procedural Complication	A070909	Oversensing		
							F26	No Health Consequences or Impact	A070909	Oversensing		
					CT3001	Implantable metal	E2403	No Clinical Signs, Symptoms or Conditions	A070909	Oversensing		
							F15	Recognized Device or Procedural Complication	A070909	Oversensing		
							F26	No Health Consequences or Impact	A070909	Oversensing		
					CT3002	Implantable synthetic polymer	E2403	No Clinical Signs, Symptoms or Conditions	A070909	Oversensing		
							F15	Recognized Device or Procedural Complication	A070909	Oversensing		
							F26	No Health Consequences or Impact	A070909	Oversensing		
					CT3006	Implantable inorganic material	E2403	No Clinical Signs, Symptoms or Conditions	A070909	Oversensing		
							F15	Recognized Device or Procedural Complication	A070909	Oversensing		
							F26	No Health Consequences or Impact	A070909	Oversensing		
					CT314	Electrically powered	E2403	No Clinical Signs, Symptoms or Conditions	A070909	Oversensing		
							F15	Recognized Device or Procedural Complication	A070909	Oversensing		
							F26	No Health Consequences or Impact	A070909	Oversensing		
					CT321	Long-term surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A070909	Oversensing		
							F15	Recognized Device or Procedural Complication	A070909	Oversensing		
							F26	No Health Consequences or Impact	A070909	Oversensing		
					CT449	Generators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070909	Oversensing		
							F15	Recognized Device or Procedural Complication	A070909	Oversensing		
							F26	No Health Consequences or Impact	A070909	Oversensing		
					CT981	Single use	E2403	No Clinical Signs, Symptoms or Conditions	A070909	Oversensing		
							F15	Recognized Device or Procedural Complication	A070909	Oversensing		
							F26	No Health Consequences or Impact	A070909	Oversensing		
					CT983	Surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A070909	Oversensing		
							F15	Recognized Device or Procedural Complication	A070909	Oversensing		
							F26	No Health Consequences or Impact	A070909	Oversensing		
9658	2019	All other reportable incidents	35852	Single-chamber implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A070909	Oversensing		
9659	2020	All other reportable incidents	47265	Dual-chamber implantable pacemaker, rate-responsive	CT141	Electrodes/Leads and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070909	Oversensing		
							F26	No Health Consequences or Impact	A070909	Oversensing		
					CT145	Cardiovascular devices	E2403	No Clinical Signs, Symptoms or Conditions	A070909	Oversensing		
							F26	No Health Consequences or Impact	A070909	Oversensing		
					CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070909	Oversensing		
							F26	No Health Consequences or Impact	A070909	Oversensing		
					CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A070909	Oversensing		
							F26	No Health Consequences or Impact	A070909	Oversensing		
					CT233	Surgical	E2403	No Clinical Signs, Symptoms or Conditions	A070909	Oversensing		
							F26	No Health Consequences or Impact	A070909	Oversensing		
					CT2406	Implantable	E2403	No Clinical Signs, Symptoms or Conditions	A070909	Oversensing		
							F26	No Health Consequences or Impact	A070909	Oversensing		
					CT2407	Active implantable	E2403	No Clinical Signs, Symptoms or Conditions	A070909	Oversensing		
							F26	No Health Consequences or Impact	A070909	Oversensing		
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A070909	Oversensing		
							F26	No Health Consequences or Impact	A070909	Oversensing		
					CT2727	Active implant devices/ancillaries	E2403	No Clinical Signs, Symptoms or Conditions	A070909	Oversensing		
							F26	No Health Consequences or Impact	A070909	Oversensing		
					CT2750	Implantable cardiac electrical stimulators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070909	Oversensing		
							F26	No Health Consequences or Impact	A070909	Oversensing		
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A070909	Oversensing		
							F26	No Health Consequences or Impact	A070909	Oversensing		
					CT2985	Active	E2403	No Clinical Signs, Symptoms or Conditions	A070909	Oversensing		
							F26	No Health Consequences or Impact	A070909	Oversensing		
					CT3001	Implantable metal	E2403	No Clinical Signs, Symptoms or Conditions	A070909	Oversensing		
							F26	No Health Consequences or Impact	A070909	Oversensing		
					CT3002	Implantable synthetic polymer	E2403	No Clinical Signs, Symptoms or Conditions	A070909	Oversensing		
							F26	No Health Consequences or Impact	A070909	Oversensing		
					CT3006	Implantable inorganic material	E2403	No Clinical Signs, Symptoms or Conditions	A070909	Oversensing		
							F26	No Health Consequences or Impact	A070909	Oversensing		
					CT314	Electrically powered	E2403	No Clinical Signs, Symptoms or Conditions	A070909	Oversensing		
							F26	No Health Consequences or Impact	A070909	Oversensing		
					CT321	Long-term surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A070909	Oversensing		

							F26	No Health Consequences or Impact	A070909	Oversensing			
					CT449	Generators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070909	Oversensing			
							F26	No Health Consequences or Impact	A070909	Oversensing			
					CT981	Single use	E2403	No Clinical Signs, Symptoms or Conditions	A070909	Oversensing			
							F26	No Health Consequences or Impact	A070909	Oversensing			
					CT983	Surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A070909	Oversensing			
							F26	No Health Consequences or Impact	A070909	Oversensing			
9660	2020	All other reportable incidents	47270	Cardiac resynchronization therapy implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A070909	Oversensing			
							F26	No Health Consequences or Impact	A070909	Oversensing			
9661	2019	All other reportable incidents	37265	Dual-chamber implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery			
9662	2020	All other reportable incidents	47270	Cardiac resynchronization therapy implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A070909	Oversensing	07/07/2020	07/07/2020	
									A0712	Pacing Problem	07/07/2020	07/07/2020	
							F14	Prolonged Episode of Care	A070909	Oversensing	07/07/2020	07/07/2020	
									A0712	Pacing Problem	07/07/2020	07/07/2020	
							F1905	Device Revision or Replacement	A070909	Oversensing	07/07/2020	07/07/2020	
									A0712	Pacing Problem	07/07/2020	07/07/2020	
							F1908	Prolonged Surgery	A070909	Oversensing	07/07/2020	07/07/2020	
									A0712	Pacing Problem	07/07/2020	07/07/2020	
9663	2020	All other reportable incidents	47267	Single-chamber implantable pacemaker, rate-responsive	CT141	Electrodes/Leads and associated devices	E0112	Dizziness	A0712	Pacing Problem	07/07/2020	07/07/2020	
							F19	Surgical Intervention	A0712	Pacing Problem	07/07/2020	07/07/2020	
							F1905	Device Revision or Replacement	A0712	Pacing Problem	07/07/2020	07/07/2020	
					CT145	Cardiovascular devices	E0112	Dizziness	A0712	Pacing Problem	07/07/2020	07/07/2020	
							F19	Surgical Intervention	A0712	Pacing Problem	07/07/2020	07/07/2020	
							F1905	Device Revision or Replacement	A0712	Pacing Problem	07/07/2020	07/07/2020	
					CT1684	Cardiac pacemakers and associated devices	E0112	Dizziness	A0712	Pacing Problem	07/07/2020	07/07/2020	
							F19	Surgical Intervention	A0712	Pacing Problem	07/07/2020	07/07/2020	
							F1905	Device Revision or Replacement	A0712	Pacing Problem	07/07/2020	07/07/2020	
					CT1685	Cardiac pacemakers	E0112	Dizziness	A0712	Pacing Problem	07/07/2020	07/07/2020	
							F19	Surgical Intervention	A0712	Pacing Problem	07/07/2020	07/07/2020	
							F1905	Device Revision or Replacement	A0712	Pacing Problem	07/07/2020	07/07/2020	
					CT233	Surgical	E0112	Dizziness	A0712	Pacing Problem	07/07/2020	07/07/2020	
							F19	Surgical Intervention	A0712	Pacing Problem	07/07/2020	07/07/2020	
							F1905	Device Revision or Replacement	A0712	Pacing Problem	07/07/2020	07/07/2020	
					CT2406	Implantable	E0112	Dizziness	A0712	Pacing Problem	07/07/2020	07/07/2020	
							F19	Surgical Intervention	A0712	Pacing Problem	07/07/2020	07/07/2020	
							F1905	Device Revision or Replacement	A0712	Pacing Problem	07/07/2020	07/07/2020	
					CT2407	Active implantable	E0112	Dizziness	A0712	Pacing Problem	07/07/2020	07/07/2020	
							F19	Surgical Intervention	A0712	Pacing Problem	07/07/2020	07/07/2020	
							F1905	Device Revision or Replacement	A0712	Pacing Problem	07/07/2020	07/07/2020	
					CT257	Implantable cardiac pacemakers	E0112	Dizziness	A0712	Pacing Problem	07/07/2020	07/07/2020	
							F19	Surgical Intervention	A0712	Pacing Problem	07/07/2020	07/07/2020	
							F1905	Device Revision or Replacement	A0712	Pacing Problem	07/07/2020	07/07/2020	
					CT2727	Active implant devices/ancillaries	E0112	Dizziness	A0712	Pacing Problem	07/07/2020	07/07/2020	
							F19	Surgical Intervention	A0712	Pacing Problem	07/07/2020	07/07/2020	
							F1905	Device Revision or Replacement	A0712	Pacing Problem	07/07/2020	07/07/2020	
					CT2750	Implantable cardiac electrical stimulators and associated devices	E0112	Dizziness	A0712	Pacing Problem	07/07/2020	07/07/2020	
							F19	Surgical Intervention	A0712	Pacing Problem	07/07/2020	07/07/2020	
							F1905	Device Revision or Replacement	A0712	Pacing Problem	07/07/2020	07/07/2020	
					CT2866	Implantable cardiac electrical stimulators	E0112	Dizziness	A0712	Pacing Problem	07/07/2020	07/07/2020	
							F19	Surgical Intervention	A0712	Pacing Problem	07/07/2020	07/07/2020	
							F1905	Device Revision or Replacement	A0712	Pacing Problem	07/07/2020	07/07/2020	
					CT2985	Active	E0112	Dizziness	A0712	Pacing Problem	07/07/2020	07/07/2020	
							F19	Surgical Intervention	A0712	Pacing Problem	07/07/2020	07/07/2020	
							F1905	Device Revision or Replacement	A0712	Pacing Problem	07/07/2020	07/07/2020	
					CT3001	Implantable metal	E0112	Dizziness	A0712	Pacing Problem	07/07/2020	07/07/2020	
							F19	Surgical Intervention	A0712	Pacing Problem	07/07/2020	07/07/2020	
							F1905	Device Revision or Replacement	A0712	Pacing Problem	07/07/2020	07/07/2020	
					CT3002	Implantable synthetic polymer	E0112	Dizziness	A0712	Pacing Problem	07/07/2020	07/07/2020	
							F19	Surgical Intervention	A0712	Pacing Problem	07/07/2020	07/07/2020	
							F1905	Device Revision or Replacement	A0712	Pacing Problem	07/07/2020	07/07/2020	
					CT3006	Implantable inorganic material	E0112	Dizziness	A0712	Pacing Problem	07/07/2020	07/07/2020	
							F19	Surgical Intervention	A0712	Pacing Problem	07/07/2020	07/07/2020	
							F1905	Device Revision or Replacement	A0712	Pacing Problem	07/07/2020	07/07/2020	
					CT314	Electrically powered	E0112	Dizziness	A0712	Pacing Problem	07/07/2020	07/07/2020	
							F19	Surgical Intervention	A0712	Pacing Problem	07/07/2020	07/07/2020	
							F1905	Device Revision or Replacement	A0712	Pacing Problem	07/07/2020	07/07/2020	
					CT321	Long-term surgical invasive	E0112	Dizziness	A0712	Pacing Problem	07/07/2020	07/07/2020	
							F19	Surgical Intervention	A0712	Pacing Problem	07/07/2020	07/07/2020	
							F1905	Device Revision or Replacement	A0712	Pacing Problem	07/07/2020	07/07/2020	
					CT449	Generators and associated devices	E0112	Dizziness	A0712	Pacing Problem	07/07/2020	07/07/2020	
							F19	Surgical Intervention	A0712	Pacing Problem	07/07/2020	07/07/2020	
							F1905	Device Revision or Replacement	A0712	Pacing Problem	07/07/2020	07/07/2020	
					CT981	Single use	E0112	Dizziness	A0712	Pacing Problem	07/07/2020	07/07/2020	
							F19	Surgical Intervention	A0712	Pacing Problem	07/07/2020	07/07/2020	
							F1905	Device Revision or Replacement	A0712	Pacing Problem	07/07/2020	07/07/2020	
					CT983	Surgical invasive	E0112	Dizziness	A0712	Pacing Problem	07/07/2020	07/07/2020	
							F19	Surgical Intervention	A0712	Pacing Problem	07/07/2020	07/07/2020	
							F1905	Device Revision or Replacement	A0712	Pacing Problem	07/07/2020	07/07/2020	
9664	2019	All other reportable incidents	47265	Dual-chamber implantable pacemaker, rate-responsive	CT141	Electrodes/Leads and associated devices	F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	23/08/2019	23/08/2019	
					CT145	Cardiovascular devices	F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	23/08/2019	23/08/2019	
					CT1684	Cardiac pacemakers and associated devices	F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	23/08/2019	23/08/2019	
					CT1685	Cardiac pacemakers	F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	23/08/2019	23/08/2019	
					CT233	Surgical	F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	23/08/2019	23/08/2019	
					CT2406	Implantable	F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	23/08/2019	23/08/2019	
					CT2407	Active implantable	F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	23/08/2019	23/08/2019	
					CT257	Implantable cardiac pacemakers	F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	23/08/2019	23/08/2019	
					CT2727	Active implant devices/ancillaries	F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	23/08/2019	23/08/2019	

					CT2750	Implantable cardiac electrical stimulators and associated devices	F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	23/08/2019	23/08/2019
					CT2866	Implantable cardiac electrical stimulators	F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	23/08/2019	23/08/2019
					CT2985	Active	F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	23/08/2019	23/08/2019
					CT3001	Implantable metal	F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	23/08/2019	23/08/2019
					CT3002	Implantable synthetic polymer	F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	23/08/2019	23/08/2019
					CT3006	Implantable inorganic material	F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	23/08/2019	23/08/2019
					CT314	Electrically powered	F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	23/08/2019	23/08/2019
					CT321	Long-term surgical invasive	F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	23/08/2019	23/08/2019
					CT449	Generators and associated devices	F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	23/08/2019	23/08/2019
					CT981	Single use	F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	23/08/2019	23/08/2019
					CT983	Surgical invasive	F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	23/08/2019	23/08/2019
9665	2019	All other reportable incidents	37265	Dual-chamber implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E060109	Tachycardia	A070909	Oversensing		
9666	2019	All other reportable incidents	35852	Single-chamber implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A070909	Oversensing		
9667	2020	All other reportable incidents	17882	Physiologic-monitoring defibrillation system	CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A07	Electrical /Electronic Property Problem		
							F27	Problem Identified During Non-Clinical Procedure	A07	Electrical /Electronic Property Problem		
9668	2019	All other reportable incidents	60789	Intracardiac pacemaker	CT130	Catheters and associated devices	F1905	Device Revision or Replacement	A0705	Battery Problem		
					CT145	Cardiovascular devices	F1905	Device Revision or Replacement	A0705	Battery Problem		
					CT1684	Cardiac pacemakers and associated devices	F1905	Device Revision or Replacement	A0705	Battery Problem		
					CT1685	Cardiac pacemakers	F1905	Device Revision or Replacement	A0705	Battery Problem		
					CT233	Surgical	F1905	Device Revision or Replacement	A0705	Battery Problem		
					CT2406	Implantable	F1905	Device Revision or Replacement	A0705	Battery Problem		
					CT2407	Active implantable	F1905	Device Revision or Replacement	A0705	Battery Problem		
					CT257	Implantable cardiac pacemakers	F1905	Device Revision or Replacement	A0705	Battery Problem		
					CT2727	Active implant devices/ancillaries	F1905	Device Revision or Replacement	A0705	Battery Problem		
					CT2750	Implantable cardiac electrical stimulators and associated devices	F1905	Device Revision or Replacement	A0705	Battery Problem		
					CT2818	Catheterization/Cannulation/Fluid path devices	F1905	Device Revision or Replacement	A0705	Battery Problem		
					CT2866	Implantable cardiac electrical stimulators	F1905	Device Revision or Replacement	A0705	Battery Problem		
					CT2985	Active	F1905	Device Revision or Replacement	A0705	Battery Problem		
					CT3001	Implantable metal	F1905	Device Revision or Replacement	A0705	Battery Problem		
					CT3002	Implantable synthetic polymer	F1905	Device Revision or Replacement	A0705	Battery Problem		
					CT3006	Implantable inorganic material	F1905	Device Revision or Replacement	A0705	Battery Problem		
					CT314	Electrically powered	F1905	Device Revision or Replacement	A0705	Battery Problem		
					CT321	Long-term surgical invasive	F1905	Device Revision or Replacement	A0705	Battery Problem		
					CT981	Single use	F1905	Device Revision or Replacement	A0705	Battery Problem		
					CT983	Surgical invasive	F1905	Device Revision or Replacement	A0705	Battery Problem		
9669	2019	All other reportable incidents	36241	Pacemaker/defibrillator lead	CT1684	Cardiac pacemakers and associated devices	F1905	Device Revision or Replacement	A072201	High Impedance		
9670	2019	All other reportable incidents	37265	Dual-chamber implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
9671	2019	All other reportable incidents	47267	Single-chamber implantable pacemaker, rate-responsive	CT141	Electrodes/Leads and associated devices	F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery		
					CT145	Cardiovascular devices	F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery		
					CT1684	Cardiac pacemakers and associated devices	F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery		
					CT1685	Cardiac pacemakers	F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery		
					CT233	Surgical	F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery		
					CT2406	Implantable	F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery		
					CT2407	Active implantable	F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery		
					CT257	Implantable cardiac pacemakers	F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery		
					CT2727	Active implant devices/ancillaries	F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery		
					CT2750	Implantable cardiac electrical stimulators and associated devices	F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery		
					CT2866	Implantable cardiac electrical stimulators	F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery		
					CT2985	Active	F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery		
					CT3001	Implantable metal	F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery		
					CT3002	Implantable synthetic polymer	F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery		
					CT3006	Implantable inorganic material	F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery		
					CT314	Electrically powered	F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery		
					CT321	Long-term surgical invasive	F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery		
					CT449	Generators and associated devices	F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery		
					CT981	Single use	F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery		
					CT983	Surgical invasive	F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery		
9672	2020	All other reportable incidents	47265	Dual-chamber implantable pacemaker, rate-responsive	CT141	Electrodes/Leads and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070101	Failure to Capture		
									A090801	Signal Artifact/Noise		
							F26	No Health Consequences or Impact	A070101	Failure to Capture		
									A090801	Signal Artifact/Noise		
					CT145	Cardiovascular devices	E2403	No Clinical Signs, Symptoms or Conditions	A070101	Failure to Capture		
									A090801	Signal Artifact/Noise		
							F26	No Health Consequences or Impact	A070101	Failure to Capture		
									A090801	Signal Artifact/Noise		
					CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070101	Failure to Capture		
									A090801	Signal Artifact/Noise		
							F26	No Health Consequences or Impact	A070101	Failure to Capture		
									A090801	Signal Artifact/Noise		
					CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A070101	Failure to Capture		
									A090801	Signal Artifact/Noise		
							F26	No Health Consequences or Impact	A070101	Failure to Capture		
									A090801	Signal Artifact/Noise		
					CT233	Surgical	E2403	No Clinical Signs, Symptoms or Conditions	A070101	Failure to Capture		
									A090801	Signal Artifact/Noise		
							F26	No Health Consequences or Impact	A070101	Failure to Capture		
									A090801	Signal Artifact/Noise		
					CT2406	Implantable	E2403	No Clinical Signs, Symptoms or Conditions	A070101	Failure to Capture		
									A090801	Signal Artifact/Noise		
							F26	No Health Consequences or Impact	A070101	Failure to Capture		
									A090801	Signal Artifact/Noise		
					CT2407	Active implantable	E2403	No Clinical Signs, Symptoms or Conditions	A070101	Failure to Capture		
									A090801	Signal Artifact/Noise		
							F26	No Health Consequences or Impact	A070101	Failure to Capture		
									A090801	Signal Artifact/Noise		

					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A070101	Failure to Capture		
									A090801	Signal Artifact/Noise		
							F26	No Health Consequences or Impact	A070101	Failure to Capture		
									A090801	Signal Artifact/Noise		
					CT2727	Active implant devices/ancillaries	E2403	No Clinical Signs, Symptoms or Conditions	A070101	Failure to Capture		
									A090801	Signal Artifact/Noise		
							F26	No Health Consequences or Impact	A070101	Failure to Capture		
									A090801	Signal Artifact/Noise		
					CT2750	Implantable cardiac electrical stimulators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070101	Failure to Capture		
									A090801	Signal Artifact/Noise		
							F26	No Health Consequences or Impact	A070101	Failure to Capture		
									A090801	Signal Artifact/Noise		
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A070101	Failure to Capture		
									A090801	Signal Artifact/Noise		
							F26	No Health Consequences or Impact	A070101	Failure to Capture		
									A090801	Signal Artifact/Noise		
					CT2985	Active	E2403	No Clinical Signs, Symptoms or Conditions	A070101	Failure to Capture		
									A090801	Signal Artifact/Noise		
							F26	No Health Consequences or Impact	A070101	Failure to Capture		
									A090801	Signal Artifact/Noise		
					CT3001	Implantable metal	E2403	No Clinical Signs, Symptoms or Conditions	A070101	Failure to Capture		
									A090801	Signal Artifact/Noise		
							F26	No Health Consequences or Impact	A070101	Failure to Capture		
									A090801	Signal Artifact/Noise		
					CT3002	Implantable synthetic polymer	E2403	No Clinical Signs, Symptoms or Conditions	A070101	Failure to Capture		
									A090801	Signal Artifact/Noise		
							F26	No Health Consequences or Impact	A070101	Failure to Capture		
									A090801	Signal Artifact/Noise		
					CT3006	Implantable inorganic material	E2403	No Clinical Signs, Symptoms or Conditions	A070101	Failure to Capture		
									A090801	Signal Artifact/Noise		
							F26	No Health Consequences or Impact	A070101	Failure to Capture		
									A090801	Signal Artifact/Noise		
					CT314	Electrically powered	E2403	No Clinical Signs, Symptoms or Conditions	A070101	Failure to Capture		
									A090801	Signal Artifact/Noise		
							F26	No Health Consequences or Impact	A070101	Failure to Capture		
									A090801	Signal Artifact/Noise		
					CT321	Long-term surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A070101	Failure to Capture		
									A090801	Signal Artifact/Noise		
							F26	No Health Consequences or Impact	A070101	Failure to Capture		
									A090801	Signal Artifact/Noise		
					CT449	Generators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070101	Failure to Capture		
									A090801	Signal Artifact/Noise		
							F26	No Health Consequences or Impact	A070101	Failure to Capture		
									A090801	Signal Artifact/Noise		
					CT981	Single use	E2403	No Clinical Signs, Symptoms or Conditions	A070101	Failure to Capture		
									A090801	Signal Artifact/Noise		
							F26	No Health Consequences or Impact	A070101	Failure to Capture		
									A090801	Signal Artifact/Noise		
					CT983	Surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A070101	Failure to Capture		
									A090801	Signal Artifact/Noise		
							F26	No Health Consequences or Impact	A070101	Failure to Capture		
									A090801	Signal Artifact/Noise		
9673	2020	All other reportable incidents	47267	Single-chamber implantable pacemaker, rate-responsive	CT141	Electrodes/Leads and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery		
							F12	Serious Injury/ Illness/ Impairment	A070504	Premature Discharge of Battery		
							F1903	Device Explantation	A070504	Premature Discharge of Battery		
					CT145	Cardiovascular devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery		
							F12	Serious Injury/ Illness/ Impairment	A070504	Premature Discharge of Battery		
							F1903	Device Explantation	A070504	Premature Discharge of Battery		
					CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery		
							F12	Serious Injury/ Illness/ Impairment	A070504	Premature Discharge of Battery		
							F1903	Device Explantation	A070504	Premature Discharge of Battery		
					CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery		
							F12	Serious Injury/ Illness/ Impairment	A070504	Premature Discharge of Battery		
							F1903	Device Explantation	A070504	Premature Discharge of Battery		
					CT233	Surgical	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery		
							F12	Serious Injury/ Illness/ Impairment	A070504	Premature Discharge of Battery		
							F1903	Device Explantation	A070504	Premature Discharge of Battery		
					CT2406	Implantable	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery		
							F12	Serious Injury/ Illness/ Impairment	A070504	Premature Discharge of Battery		
							F1903	Device Explantation	A070504	Premature Discharge of Battery		
					CT2407	Active implantable	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery		
							F12	Serious Injury/ Illness/ Impairment	A070504	Premature Discharge of Battery		
							F1903	Device Explantation	A070504	Premature Discharge of Battery		
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery		
							F12	Serious Injury/ Illness/ Impairment	A070504	Premature Discharge of Battery		
							F1903	Device Explantation	A070504	Premature Discharge of Battery		
					CT2727	Active implant devices/ancillaries	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery		
							F12	Serious Injury/ Illness/ Impairment	A070504	Premature Discharge of Battery		
							F1903	Device Explantation	A070504	Premature Discharge of Battery		

					CT2750	Implantable cardiac electrical stimulators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery		
							F12	Serious Injury/ Illness/ Impairment	A070504	Premature Discharge of Battery		
							F1903	Device Explantation	A070504	Premature Discharge of Battery		
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery		
							F12	Serious Injury/ Illness/ Impairment	A070504	Premature Discharge of Battery		
							F1903	Device Explantation	A070504	Premature Discharge of Battery		
					CT2985	Active	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery		
							F12	Serious Injury/ Illness/ Impairment	A070504	Premature Discharge of Battery		
							F1903	Device Explantation	A070504	Premature Discharge of Battery		
					CT3001	Implantable metal	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery		
							F12	Serious Injury/ Illness/ Impairment	A070504	Premature Discharge of Battery		
							F1903	Device Explantation	A070504	Premature Discharge of Battery		
					CT3002	Implantable synthetic polymer	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery		
							F12	Serious Injury/ Illness/ Impairment	A070504	Premature Discharge of Battery		
							F1903	Device Explantation	A070504	Premature Discharge of Battery		
					CT3006	Implantable inorganic material	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery		
							F12	Serious Injury/ Illness/ Impairment	A070504	Premature Discharge of Battery		
							F1903	Device Explantation	A070504	Premature Discharge of Battery		
					CT314	Electrically powered	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery		
							F12	Serious Injury/ Illness/ Impairment	A070504	Premature Discharge of Battery		
							F1903	Device Explantation	A070504	Premature Discharge of Battery		
					CT321	Long-term surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery		
							F12	Serious Injury/ Illness/ Impairment	A070504	Premature Discharge of Battery		
							F1903	Device Explantation	A070504	Premature Discharge of Battery		
					CT449	Generators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery		
							F12	Serious Injury/ Illness/ Impairment	A070504	Premature Discharge of Battery		
							F1903	Device Explantation	A070504	Premature Discharge of Battery		
					CT981	Single use	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery		
							F12	Serious Injury/ Illness/ Impairment	A070504	Premature Discharge of Battery		
							F1903	Device Explantation	A070504	Premature Discharge of Battery		
					CT983	Surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery		
							F12	Serious Injury/ Illness/ Impairment	A070504	Premature Discharge of Battery		
							F1903	Device Explantation	A070504	Premature Discharge of Battery		
9674	2019	All other reportable incidents	35852	Single-chamber implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A25	No Apparent Adverse Event	09/09/2019	
9675	2019	All other reportable incidents	36241	Pacemaker/defibrillator lead	CT1684	Cardiac pacemakers and associated devices	F1905	Device Revision or Replacement	A0510	Retraction Problem	19/08/2019	
9676	2019	All other reportable incidents	36241	Pacemaker/defibrillator lead	CT1684	Cardiac pacemakers and associated devices	F1905	Device Revision or Replacement	A04	Material Integrity Problem	20/08/2019	
9677	2019	All other reportable incidents	36241	Pacemaker/defibrillator lead	CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A	Annex A		
9678	2019	All other reportable incidents	37265	Dual-chamber implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
9679	2020	All other reportable incidents	36241	Pacemaker/defibrillator lead	CT1684	Cardiac pacemakers and associated devices	E0623	Unspecified Heart Problem	A051201	Device Dislodged or Dislocated	17/07/2020	17/07/2020
									A070101	Failure to Capture	17/07/2020	17/07/2020
							E233001	Chest Pain	A051201	Device Dislodged or Dislocated	17/07/2020	17/07/2020
									A070101	Failure to Capture	17/07/2020	17/07/2020
							F12	Serious Injury/ Illness/ Impairment	A051201	Device Dislodged or Dislocated	17/07/2020	17/07/2020
									A070101	Failure to Capture	17/07/2020	17/07/2020
							F1905	Device Revision or Replacement	A051201	Device Dislodged or Dislocated	17/07/2020	17/07/2020
									A070101	Failure to Capture	17/07/2020	17/07/2020
9680	2020	All other reportable incidents	47270	Cardiac resynchronization therapy implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A070101	Failure to Capture		
							F26	No Health Consequences or Impact	A070101	Failure to Capture		
9681	2020	All other reportable incidents	37265	Dual-chamber implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A070909	Oversensing		
							F26	No Health Consequences or Impact	A070909	Oversensing		
9682	2020	All other reportable incidents	36241	Pacemaker/defibrillator lead	CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070909	Oversensing		
									A072202	Low Impedance		
									A090801	Signal Artifact/Noise		
							F26	No Health Consequences or Impact	A070909	Oversensing		
									A072202	Low Impedance		
									A090801	Signal Artifact/Noise		
9683	2020	All other reportable incidents	36241	Pacemaker/defibrillator lead	CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070909	Oversensing		
									A0722	Impedance Problem		
									A090801	Signal Artifact/Noise		
							F26	No Health Consequences or Impact	A070909	Oversensing		
									A0722	Impedance Problem		
									A090801	Signal Artifact/Noise		
9684	2020	All other reportable incidents	36241	Pacemaker/defibrillator lead	CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A05	Mechanical Problem		
									A070102	High Capture Threshold		
							F26	No Health Consequences or Impact	A05	Mechanical Problem		
									A070102	High Capture Threshold		
9685	2019	All other reportable incidents	17882	Physiologic-monitoring defibrillation system	CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A1303	Intermittent Communication Failure		
9686	2019	All other reportable incidents	36241	Pacemaker/defibrillator lead	CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A05	Mechanical Problem	20/08/2019	20/08/2019
9687	2019	All other reportable incidents	47265	Dual-chamber implantable pacemaker, rate-responsive	CT141	Electrodes/Leads and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
					CT145	Cardiovascular devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
					CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
					CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
					CT233	Surgical	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
					CT2406	Implantable	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
					CT2407	Active implantable	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
					CT2727	Active implant devices/ancillaries	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		

					CT2750	Implantable cardiac electrical stimulators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
					CT2985	Active	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
					CT3001	Implantable metal	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
					CT3002	Implantable synthetic polymer	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
					CT3006	Implantable inorganic material	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
					CT314	Electrically powered	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
					CT321	Long-term surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
					CT449	Generators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
					CT981	Single use	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
					CT983	Surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
9688	2019	All other reportable incidents	36241	Pacemaker/defibrillator lead	CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A072201	High Impedance		
9689	2020	All other reportable incidents	37265	Dual-chamber implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E011903	Syncope/Fainting	A0712	Pacing Problem		
									A071202	Inaccurate Synchronization		
									A071302	Inappropriate/Inadequate Shock/Stimulation		
									A0909	Unexpected Therapeutic Results		
							E060109	Tachycardia	A0712	Pacing Problem		
									A071202	Inaccurate Synchronization		
									A071302	Inappropriate/Inadequate Shock/Stimulation		
									A0909	Unexpected Therapeutic Results		
							E2104	Electric Shock	A0712	Pacing Problem		
									A071202	Inaccurate Synchronization		
									A071302	Inappropriate/Inadequate Shock/Stimulation		
									A0909	Unexpected Therapeutic Results		
							F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A0712	Pacing Problem		
									A071202	Inaccurate Synchronization		
									A071302	Inappropriate/Inadequate Shock/Stimulation		
									A0909	Unexpected Therapeutic Results		
							F14	Prolonged Episode of Care	A0712	Pacing Problem		
									A071202	Inaccurate Synchronization		
									A071302	Inappropriate/Inadequate Shock/Stimulation		
									A0909	Unexpected Therapeutic Results		
							F23	Unexpected Medical Intervention	A0712	Pacing Problem		
									A071202	Inaccurate Synchronization		
									A071302	Inappropriate/Inadequate Shock/Stimulation		
									A0909	Unexpected Therapeutic Results		
9690	2020	All other reportable incidents	35852	Single-chamber implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	17/08/2020	17/08/2020
							F19	Surgical Intervention	A070504	Premature Discharge of Battery	17/08/2020	17/08/2020
9691	2020	All other reportable incidents	37265	Dual-chamber implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A070909	Oversensing		
							F26	No Health Consequences or Impact	A070909	Oversensing		
9692	2020	All other reportable incidents	36241	Pacemaker/defibrillator lead	CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A0203	Defective Device	10/07/2020	10/07/2020
							F1908	Prolonged Surgery	A0203	Defective Device	10/07/2020	10/07/2020
9693	2020	All other reportable incidents	47270	Cardiac resynchronization therapy implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A071208	Pacing Intermittently		
							F14	Prolonged Episode of Care	A071208	Pacing Intermittently		
							F26	No Health Consequences or Impact	A071208	Pacing Intermittently		
9694	2020	All other reportable incidents	47266	Single-chamber implantable pacemaker, fixed-rate/demand	CT141	Electrodes/Leads and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A0722	Impedance Problem		
							F26	No Health Consequences or Impact	A0722	Impedance Problem		
					CT145	Cardiovascular devices	E2403	No Clinical Signs, Symptoms or Conditions	A0722	Impedance Problem		
							F26	No Health Consequences or Impact	A0722	Impedance Problem		
					CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A0722	Impedance Problem		
							F26	No Health Consequences or Impact	A0722	Impedance Problem		
					CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A0722	Impedance Problem		
							F26	No Health Consequences or Impact	A0722	Impedance Problem		
					CT233	Surgical	E2403	No Clinical Signs, Symptoms or Conditions	A0722	Impedance Problem		
							F26	No Health Consequences or Impact	A0722	Impedance Problem		
					CT2406	Implantable	E2403	No Clinical Signs, Symptoms or Conditions	A0722	Impedance Problem		
							F26	No Health Consequences or Impact	A0722	Impedance Problem		
					CT2407	Active implantable	E2403	No Clinical Signs, Symptoms or Conditions	A0722	Impedance Problem		
							F26	No Health Consequences or Impact	A0722	Impedance Problem		
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A0722	Impedance Problem		
							F26	No Health Consequences or Impact	A0722	Impedance Problem		
					CT2727	Active implant devices/ancillaries	E2403	No Clinical Signs, Symptoms or Conditions	A0722	Impedance Problem		
							F26	No Health Consequences or Impact	A0722	Impedance Problem		
					CT2750	Implantable cardiac electrical stimulators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A0722	Impedance Problem		
							F26	No Health Consequences or Impact	A0722	Impedance Problem		
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A0722	Impedance Problem		
							F26	No Health Consequences or Impact	A0722	Impedance Problem		
					CT2985	Active	E2403	No Clinical Signs, Symptoms or Conditions	A0722	Impedance Problem		
							F26	No Health Consequences or Impact	A0722	Impedance Problem		
					CT3001	Implantable metal	E2403	No Clinical Signs, Symptoms or Conditions	A0722	Impedance Problem		
							F26	No Health Consequences or Impact	A0722	Impedance Problem		
					CT3002	Implantable synthetic polymer	E2403	No Clinical Signs, Symptoms or Conditions	A0722	Impedance Problem		
							F26	No Health Consequences or Impact	A0722	Impedance Problem		
					CT3006	Implantable inorganic material	E2403	No Clinical Signs, Symptoms or Conditions	A0722	Impedance Problem		
							F26	No Health Consequences or Impact	A0722	Impedance Problem		
					CT314	Electrically powered	E2403	No Clinical Signs, Symptoms or Conditions	A0722	Impedance Problem		
							F26	No Health Consequences or Impact	A0722	Impedance Problem		
					CT321	Long-term surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A0722	Impedance Problem		
							F26	No Health Consequences or Impact	A0722	Impedance Problem		
					CT449	Generators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A0722	Impedance Problem		
							F26	No Health Consequences or Impact	A0722	Impedance Problem		
					CT981	Single use	E2403	No Clinical Signs, Symptoms or Conditions	A0722	Impedance Problem		
							F26	No Health Consequences or Impact	A0722	Impedance Problem		
					CT983	Surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A0722	Impedance Problem		
							F26	No Health Consequences or Impact	A0722	Impedance Problem		
9695	2020	All other reportable incidents	37265	Dual-chamber implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A0712	Pacing Problem		
							F15	Recognized Device or Procedural Complication	A0712	Pacing Problem		

9696	2020	All other reportable incidents	35852	Single-chamber implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A0712	Pacing Problem		
							F2303	Medication Required	A0712	Pacing Problem		
9697	2020	All other reportable incidents	36241	Pacemaker/defibrillator lead	CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070909	Oversensing		
							F26	No Health Consequences or Impact	A070909	Oversensing		
9698	2019	All other reportable incidents	35852	Single-chamber implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A072201	High Impedance		
9699	2019	All other reportable incidents	35223	Endocardial/interventricular septal pacing lead	CT1684	Cardiac pacemakers and associated devices	F1905	Device Revision or Replacement	A040101	Fracture		
9700	2019	All other reportable incidents	60910	Coronary venous pacing lead	CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A0722	Impedance Problem		
9701	2020	All other reportable incidents	35852	Single-chamber implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	05/08/2020	05/08/2020
							F1901	Additional Surgery	A070504	Premature Discharge of Battery	05/08/2020	05/08/2020
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	05/08/2020	05/08/2020
9702	2019	All other reportable incidents	35852	Single-chamber implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A070910	Undersensing		
9703	2020	All other reportable incidents	35223	Endocardial/interventricular septal pacing lead	CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070910	Undersensing		
									A072202	Low Impedance		
									A25	No Apparent Adverse Event		
							F1905	Device Revision or Replacement	A070910	Undersensing		
									A072202	Low Impedance		
									A25	No Apparent Adverse Event		
							F26	No Health Consequences or Impact	A070910	Undersensing		
									A072202	Low Impedance		
									A25	No Apparent Adverse Event		
9704	2019	Death	37265	Dual-chamber implantable defibrillator	CT2866	Implantable cardiac electrical stimulators			A0908	Incorrect, Inadequate or Imprecise Result or Readings	06/09/2019	06/09/2019
9705	2020	All other reportable incidents	37265	Dual-chamber implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A070910	Undersensing	23/07/2020	23/07/2020
									A072202	Low Impedance	23/07/2020	23/07/2020
									A25	No Apparent Adverse Event	23/07/2020	23/07/2020
							F26	No Health Consequences or Impact	A070910	Undersensing	23/07/2020	23/07/2020
									A072202	Low Impedance	23/07/2020	23/07/2020
									A25	No Apparent Adverse Event	23/07/2020	23/07/2020
9706	2019	All other reportable incidents	35852	Single-chamber implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A070909	Oversensing		
9707	2019	All other reportable incidents	36241	Pacemaker/defibrillator lead	CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070909	Oversensing		
9708	2020	Death	47270	Cardiac resynchronization therapy implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E2401	Insufficient Information	A0709	Device Sensing Problem	01/07/2020	31/07/2020
							F02	Death	A0709	Device Sensing Problem	01/07/2020	31/07/2020
							F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A0709	Device Sensing Problem	01/07/2020	31/07/2020
9709	2020	All other reportable incidents	36241	Pacemaker/defibrillator lead	CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A072201	High Impedance		
							F26	No Health Consequences or Impact	A072201	High Impedance		
9710	2020	All other reportable incidents	47270	Cardiac resynchronization therapy implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A072201	High Impedance		
							F26	No Health Consequences or Impact	A072201	High Impedance		
9711	2019	All other reportable incidents	35852	Single-chamber implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E2104	Electric Shock	A071302	Inappropriate/Inadequate Shock/Stimulation		
							E2330	Pain	A071302	Inappropriate/Inadequate Shock/Stimulation		
							F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A071302	Inappropriate/Inadequate Shock/Stimulation		
							F1905	Device Revision or Replacement	A071302	Inappropriate/Inadequate Shock/Stimulation		
9712	2019	All other reportable incidents	36241	Pacemaker/defibrillator lead	CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A072201	High Impedance		
9713	2019	All other reportable incidents	35852	Single-chamber implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	F1905	Device Revision or Replacement	A26	Insufficient Device Problem Information		
9714	2019	All other reportable incidents	47270	Cardiac resynchronization therapy implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
9715	2019	All other reportable incidents	47265	Dual-chamber implantable pacemaker, rate-responsive	CT141	Electrodes/Leads and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	21/08/2019	
					CT145	Cardiovascular devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	21/08/2019	
					CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	21/08/2019	
					CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	21/08/2019	
					CT233	Surgical	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	21/08/2019	
					CT2406	Implantable	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	21/08/2019	
					CT2407	Active implantable	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	21/08/2019	
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	21/08/2019	
					CT2727	Active implant devices/ancillaries	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	21/08/2019	
					CT2750	Implantable cardiac electrical stimulators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	21/08/2019	
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	21/08/2019	
					CT2985	Active	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	21/08/2019	
					CT3001	Implantable metal	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	21/08/2019	
					CT3002	Implantable synthetic polymer	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	21/08/2019	
					CT3006	Implantable inorganic material	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	21/08/2019	
					CT314	Electrically powered	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	21/08/2019	
					CT321	Long-term surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	21/08/2019	
					CT449	Generators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	21/08/2019	
					CT981	Single use	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	21/08/2019	
					CT983	Surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	21/08/2019	
9716	2019	All other reportable incidents	35852	Single-chamber implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
9717	2020	All other reportable incidents	37265	Dual-chamber implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	28/07/2020	28/07/2020
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	28/07/2020	28/07/2020
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery	28/07/2020	28/07/2020
9718	2019	All other reportable incidents	47263	Cardiac resynchronization therapy implantable pacemaker	CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A071202	Inaccurate Synchronization		
					CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A071202	Inaccurate Synchronization		
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A071202	Inaccurate Synchronization		
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A071202	Inaccurate Synchronization		
9719	2019	All other reportable incidents	17882	Physiologic-monitoring defibrillation system	CT1684	Cardiac pacemakers and associated devices	F05	Delay to Treatment/ Therapy	A071301	Failure to Deliver Shock/Stimulation		
9720	2019	Death	17882	Physiologic-monitoring defibrillation system	CT1684	Cardiac pacemakers and associated devices	F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A071301	Failure to Deliver Shock/Stimulation		
9721	2019	All other reportable incidents	35223	Endocardial/interventricular septal pacing lead	CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A072202	Low Impedance		
9722	2019	All other reportable incidents	35223	Endocardial/interventricular septal pacing lead	CT1684	Cardiac pacemakers and associated devices	F1905	Device Revision or Replacement	A	Annex A		
9723	2019	All other reportable incidents	35223	Endocardial/interventricular septal pacing lead	CT1684	Cardiac pacemakers and associated devices	F1905	Device Revision or Replacement	A072201	High Impedance		
9724	2020	Death	47270	Cardiac resynchronization therapy implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E2401	Insufficient Information	A27	Appropriate Device Problem Term/Code not Available	01/01/2018	31/12/2018
							F02	Death	A27	Appropriate Device Problem Term/Code not Available	01/01/2018	31/12/2018
							F1905	Device Revision or Replacement	A27	Appropriate Device Problem Term/Code not Available	01/01/2018	31/12/2018
9725	2020	All other reportable incidents	35223	Endocardial/interventricular septal pacing lead	CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A0401	Break		

									A070909	Oversensing		
							F26	No Health Consequences or Impact	A0401	Break		
									A070909	Oversensing		
9726	2020	All other reportable incidents	35852	Single-chamber implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	19/10/2020	19/10/2020
							F19	Surgical Intervention	A070504	Premature Discharge of Battery	19/10/2020	19/10/2020
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	19/10/2020	19/10/2020
9727	2020	All other reportable incidents	35852	Single-chamber implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E2104	Electric Shock	A071302	Inappropriate/Inadequate Shock/Stimulation		
							F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A071302	Inappropriate/Inadequate Shock/Stimulation		
							F26	No Health Consequences or Impact	A071302	Inappropriate/Inadequate Shock/Stimulation		
9728	2020	All other reportable incidents	35224	External pacemaker, epicardial pacing	CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A0709	Device Sensing Problem		
							F27	Problem Identified During Non-Clinical Procedure	A0709	Device Sensing Problem		
					CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A0709	Device Sensing Problem		
							F27	Problem Identified During Non-Clinical Procedure	A0709	Device Sensing Problem		
9729	2020	All other reportable incidents	35224	External pacemaker, epicardial pacing	CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A0709	Device Sensing Problem		
							F27	Problem Identified During Non-Clinical Procedure	A0709	Device Sensing Problem		
					CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A0709	Device Sensing Problem		
							F27	Problem Identified During Non-Clinical Procedure	A0709	Device Sensing Problem		
9730	2019	All other reportable incidents	17882	Physiologic-monitoring defibrillation system	CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070801	Complete Loss of Power		
9731	2020	All other reportable incidents	37265	Dual-chamber implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	04/08/2020	04/08/2020
							F1901	Additional Surgery	A070504	Premature Discharge of Battery	04/08/2020	04/08/2020
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	04/08/2020	04/08/2020
9732	2020	All other reportable incidents	35223	Endocardial/interventricular septal pacing lead	CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A040101	Fracture		
									A070102	High Capture Threshold		
									A072201	High Impedance		
							F26	No Health Consequences or Impact	A040101	Fracture		
									A070102	High Capture Threshold		
									A072201	High Impedance		
9733	2020	All other reportable incidents	37265	Dual-chamber implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
9734	2020	All other reportable incidents	35852	Single-chamber implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	05/08/2020	05/08/2020
							F19	Surgical Intervention	A070504	Premature Discharge of Battery	05/08/2020	05/08/2020
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	05/08/2020	05/08/2020
9735	2020	All other reportable incidents	47263	Cardiac resynchronization therapy implantable pacemaker	CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070909	Oversensing		
									A090804	False Positive Result		
							F26	No Health Consequences or Impact	A070909	Oversensing		
									A090804	False Positive Result		
					CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A070909	Oversensing		
									A090804	False Positive Result		
							F26	No Health Consequences or Impact	A070909	Oversensing		
									A090804	False Positive Result		
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A070909	Oversensing		
									A090804	False Positive Result		
							F26	No Health Consequences or Impact	A070909	Oversensing		
									A090804	False Positive Result		
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A070909	Oversensing		
									A090804	False Positive Result		
							F26	No Health Consequences or Impact	A070909	Oversensing		
									A090804	False Positive Result		
9736	2019	All other reportable incidents	35223	Endocardial/interventricular septal pacing lead	CT1684	Cardiac pacemakers and associated devices	F21	Unexpected Deterioration	A070101	Failure to Capture		
9737	2019	All other reportable incidents	47270	Cardiac resynchronization therapy implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	11/09/2019	
9738	2019	All other reportable incidents	47265	Dual-chamber implantable pacemaker, rate-responsive	CT141	Electrodes/Leads and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	30/10/2019	30/10/2019
					CT145	Cardiovascular devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	30/10/2019	30/10/2019
					CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	30/10/2019	30/10/2019
					CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	30/10/2019	30/10/2019
					CT233	Surgical	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	30/10/2019	30/10/2019
					CT2406	Implantable	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	30/10/2019	30/10/2019
					CT2407	Active implantable	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	30/10/2019	30/10/2019
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	30/10/2019	30/10/2019
					CT2727	Active implant devices/ancillaries	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	30/10/2019	30/10/2019
					CT2750	Implantable cardiac electrical stimulators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	30/10/2019	30/10/2019
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	30/10/2019	30/10/2019
					CT2985	Active	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	30/10/2019	30/10/2019
					CT3001	Implantable metal	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	30/10/2019	30/10/2019
					CT3002	Implantable synthetic polymer	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	30/10/2019	30/10/2019
					CT3006	Implantable inorganic material	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	30/10/2019	30/10/2019
					CT314	Electrically powered	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	30/10/2019	30/10/2019
					CT321	Long-term surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	30/10/2019	30/10/2019
					CT449	Generators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	30/10/2019	30/10/2019
					CT981	Single use	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	30/10/2019	30/10/2019
					CT983	Surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	30/10/2019	30/10/2019
9739	2019	All other reportable incidents	37265	Dual-chamber implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	18/09/2019	
9740	2019	All other reportable incidents	35852	Single-chamber implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A071302	Inappropriate/Inadequate Shock/Stimulation		
9741	2019	All other reportable incidents	37265	Dual-chamber implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
9742	2020	All other reportable incidents	37265	Dual-chamber implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
9743	2020	All other reportable incidents	47267	Single-chamber implantable pacemaker, rate-responsive	CT141	Electrodes/Leads and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	13/08/2020	13/08/2020
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	13/08/2020	13/08/2020
							F12	Serious Injury/ Illness/ Impairment	A070504	Premature Discharge of Battery	13/08/2020	13/08/2020
							F19	Surgical Intervention	A070504	Premature Discharge of Battery	13/08/2020	13/08/2020
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	13/08/2020	13/08/2020
					CT145	Cardiovascular devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	13/08/2020	13/08/2020
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	13/08/2020	13/08/2020
							F12	Serious Injury/ Illness/ Impairment	A070504	Premature Discharge of Battery	13/08/2020	13/08/2020
							F19	Surgical Intervention	A070504	Premature Discharge of Battery	13/08/2020	13/08/2020
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	13/08/2020	13/08/2020

					CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	13/08/2020	13/08/2020
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	13/08/2020	13/08/2020
							F12	Serious Injury/ Illness/ Impairment	A070504	Premature Discharge of Battery	13/08/2020	13/08/2020
							F19	Surgical Intervention	A070504	Premature Discharge of Battery	13/08/2020	13/08/2020
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	13/08/2020	13/08/2020
					CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	13/08/2020	13/08/2020
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	13/08/2020	13/08/2020
							F12	Serious Injury/ Illness/ Impairment	A070504	Premature Discharge of Battery	13/08/2020	13/08/2020
							F19	Surgical Intervention	A070504	Premature Discharge of Battery	13/08/2020	13/08/2020
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	13/08/2020	13/08/2020
					CT233	Surgical	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	13/08/2020	13/08/2020
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	13/08/2020	13/08/2020
							F12	Serious Injury/ Illness/ Impairment	A070504	Premature Discharge of Battery	13/08/2020	13/08/2020
							F19	Surgical Intervention	A070504	Premature Discharge of Battery	13/08/2020	13/08/2020
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	13/08/2020	13/08/2020
					CT2406	Implantable	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	13/08/2020	13/08/2020
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	13/08/2020	13/08/2020
							F12	Serious Injury/ Illness/ Impairment	A070504	Premature Discharge of Battery	13/08/2020	13/08/2020
							F19	Surgical Intervention	A070504	Premature Discharge of Battery	13/08/2020	13/08/2020
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	13/08/2020	13/08/2020
					CT2407	Active implantable	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	13/08/2020	13/08/2020
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	13/08/2020	13/08/2020
							F12	Serious Injury/ Illness/ Impairment	A070504	Premature Discharge of Battery	13/08/2020	13/08/2020
							F19	Surgical Intervention	A070504	Premature Discharge of Battery	13/08/2020	13/08/2020
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	13/08/2020	13/08/2020
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	13/08/2020	13/08/2020
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	13/08/2020	13/08/2020
							F12	Serious Injury/ Illness/ Impairment	A070504	Premature Discharge of Battery	13/08/2020	13/08/2020
							F19	Surgical Intervention	A070504	Premature Discharge of Battery	13/08/2020	13/08/2020
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	13/08/2020	13/08/2020
					CT2727	Active implant devices/ancillaries	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	13/08/2020	13/08/2020
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	13/08/2020	13/08/2020
							F12	Serious Injury/ Illness/ Impairment	A070504	Premature Discharge of Battery	13/08/2020	13/08/2020
							F19	Surgical Intervention	A070504	Premature Discharge of Battery	13/08/2020	13/08/2020
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	13/08/2020	13/08/2020
					CT2750	Implantable cardiac electrical stimulators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	13/08/2020	13/08/2020
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	13/08/2020	13/08/2020
							F12	Serious Injury/ Illness/ Impairment	A070504	Premature Discharge of Battery	13/08/2020	13/08/2020
							F19	Surgical Intervention	A070504	Premature Discharge of Battery	13/08/2020	13/08/2020
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	13/08/2020	13/08/2020
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	13/08/2020	13/08/2020
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	13/08/2020	13/08/2020
							F12	Serious Injury/ Illness/ Impairment	A070504	Premature Discharge of Battery	13/08/2020	13/08/2020
							F19	Surgical Intervention	A070504	Premature Discharge of Battery	13/08/2020	13/08/2020
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	13/08/2020	13/08/2020
					CT2985	Active	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	13/08/2020	13/08/2020
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	13/08/2020	13/08/2020
							F12	Serious Injury/ Illness/ Impairment	A070504	Premature Discharge of Battery	13/08/2020	13/08/2020
							F19	Surgical Intervention	A070504	Premature Discharge of Battery	13/08/2020	13/08/2020
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	13/08/2020	13/08/2020
					CT3001	Implantable metal	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	13/08/2020	13/08/2020
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	13/08/2020	13/08/2020
							F12	Serious Injury/ Illness/ Impairment	A070504	Premature Discharge of Battery	13/08/2020	13/08/2020
							F19	Surgical Intervention	A070504	Premature Discharge of Battery	13/08/2020	13/08/2020
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	13/08/2020	13/08/2020
					CT3002	Implantable synthetic polymer	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	13/08/2020	13/08/2020
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	13/08/2020	13/08/2020
							F12	Serious Injury/ Illness/ Impairment	A070504	Premature Discharge of Battery	13/08/2020	13/08/2020
							F19	Surgical Intervention	A070504	Premature Discharge of Battery	13/08/2020	13/08/2020
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	13/08/2020	13/08/2020
					CT3006	Implantable inorganic material	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	13/08/2020	13/08/2020
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	13/08/2020	13/08/2020
							F12	Serious Injury/ Illness/ Impairment	A070504	Premature Discharge of Battery	13/08/2020	13/08/2020
							F19	Surgical Intervention	A070504	Premature Discharge of Battery	13/08/2020	13/08/2020
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	13/08/2020	13/08/2020
					CT314	Electrically powered	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	13/08/2020	13/08/2020
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	13/08/2020	13/08/2020
							F12	Serious Injury/ Illness/ Impairment	A070504	Premature Discharge of Battery	13/08/2020	13/08/2020
							F19	Surgical Intervention	A070504	Premature Discharge of Battery	13/08/2020	13/08/2020
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	13/08/2020	13/08/2020
					CT321	Long-term surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	13/08/2020	13/08/2020
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	13/08/2020	13/08/2020
							F12	Serious Injury/ Illness/ Impairment	A070504	Premature Discharge of Battery	13/08/2020	13/08/2020
							F19	Surgical Intervention	A070504	Premature Discharge of Battery	13/08/2020	13/08/2020
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	13/08/2020	13/08/2020
					CT449	Generators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	13/08/2020	13/08/2020
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	13/08/2020	13/08/2020
							F12	Serious Injury/ Illness/ Impairment	A070504	Premature Discharge of Battery	13/08/2020	13/08/2020
							F19	Surgical Intervention	A070504	Premature Discharge of Battery	13/08/2020	13/08/2020
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	13/08/2020	13/08/2020
					CT981	Single use	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	13/08/2020	13/08/2020
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	13/08/2020	13/08/2020
							F12	Serious Injury/ Illness/ Impairment	A070504	Premature Discharge of Battery	13/08/2020	13/08/2020
							F19	Surgical Intervention	A070504	Premature Discharge of Battery	13/08/2020	13/08/2020
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	13/08/2020	13/08/2020
					CT983	Surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	13/08/2020	13/08/2020
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	13/08/2020	13/08/2020
							F12	Serious Injury/ Illness/ Impairment	A070504	Premature Discharge of Battery	13/08/2020	13/08/2020

							F19	Surgical Intervention	A070504	Premature Discharge of Battery	13/08/2020	13/08/2020
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	13/08/2020	13/08/2020
9744	2019	All other reportable incidents	35223	Endocardial/interventricular septal pacing lead	CT1684	Cardiac pacemakers and associated devices	F1905	Device Revision or Replacement	A010402	Migration		
9745	2019	All other reportable incidents	35852	Single-chamber implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	F10	Inadequate/inappropriate Treatment or Diagnostic Exposure	A071205	Pacemaker Found in Back-Up Mode		
9746	2020	All other reportable incidents	37265	Dual-chamber implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F24	Insufficient Health Impact Information	A070504	Premature Discharge of Battery		
9747	2020	All other reportable incidents	36241	Pacemaker/defibrillator lead	CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070101	Failure to Capture		
									A070102	High Capture Threshold		
							F26	No Health Consequences or Impact	A070101	Failure to Capture		
									A070102	High Capture Threshold		
9748	2019	All other reportable incidents	47270	Cardiac resynchronization therapy implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A072201	High Impedance		
9749	2019	All other reportable incidents	47265	Dual-chamber implantable pacemaker, rate-responsive	CT141	Electrodes/Leads and associated devices	E2007	Fall	A070101	Failure to Capture		
					CT145	Cardiovascular devices	E2007	Fall	A070101	Failure to Capture		
					CT1684	Cardiac pacemakers and associated devices	E2007	Fall	A070101	Failure to Capture		
					CT1685	Cardiac pacemakers	E2007	Fall	A070101	Failure to Capture		
					CT233	Surgical	E2007	Fall	A070101	Failure to Capture		
					CT2406	Implantable	E2007	Fall	A070101	Failure to Capture		
					CT2407	Active implantable	E2007	Fall	A070101	Failure to Capture		
					CT257	Implantable cardiac pacemakers	E2007	Fall	A070101	Failure to Capture		
					CT2727	Active implant devices/ancillaries	E2007	Fall	A070101	Failure to Capture		
					CT2750	Implantable cardiac electrical stimulators and associated devices	E2007	Fall	A070101	Failure to Capture		
					CT2866	Implantable cardiac electrical stimulators	E2007	Fall	A070101	Failure to Capture		
					CT2985	Active	E2007	Fall	A070101	Failure to Capture		
					CT3001	Implantable metal	E2007	Fall	A070101	Failure to Capture		
					CT3002	Implantable synthetic polymer	E2007	Fall	A070101	Failure to Capture		
					CT3006	Implantable inorganic material	E2007	Fall	A070101	Failure to Capture		
					CT314	Electrically powered	E2007	Fall	A070101	Failure to Capture		
					CT321	Long-term surgical invasive	E2007	Fall	A070101	Failure to Capture		
					CT449	Generators and associated devices	E2007	Fall	A070101	Failure to Capture		
					CT981	Single use	E2007	Fall	A070101	Failure to Capture		
					CT983	Surgical invasive	E2007	Fall	A070101	Failure to Capture		
9750	2019	All other reportable incidents	47265	Dual-chamber implantable pacemaker, rate-responsive	CT141	Electrodes/Leads and associated devices	F1905	Device Revision or Replacement	A071205	Pacemaker Found in Back-Up Mode	30/08/2019	
					CT145	Cardiovascular devices	F1905	Device Revision or Replacement	A071205	Pacemaker Found in Back-Up Mode	30/08/2019	
					CT1684	Cardiac pacemakers and associated devices	F1905	Device Revision or Replacement	A071205	Pacemaker Found in Back-Up Mode	30/08/2019	
					CT1685	Cardiac pacemakers	F1905	Device Revision or Replacement	A071205	Pacemaker Found in Back-Up Mode	30/08/2019	
					CT233	Surgical	F1905	Device Revision or Replacement	A071205	Pacemaker Found in Back-Up Mode	30/08/2019	
					CT2406	Implantable	F1905	Device Revision or Replacement	A071205	Pacemaker Found in Back-Up Mode	30/08/2019	
					CT2407	Active implantable	F1905	Device Revision or Replacement	A071205	Pacemaker Found in Back-Up Mode	30/08/2019	
					CT257	Implantable cardiac pacemakers	F1905	Device Revision or Replacement	A071205	Pacemaker Found in Back-Up Mode	30/08/2019	
					CT2727	Active implant devices/ancillaries	F1905	Device Revision or Replacement	A071205	Pacemaker Found in Back-Up Mode	30/08/2019	
					CT2750	Implantable cardiac electrical stimulators and associated devices	F1905	Device Revision or Replacement	A071205	Pacemaker Found in Back-Up Mode	30/08/2019	
					CT2866	Implantable cardiac electrical stimulators	F1905	Device Revision or Replacement	A071205	Pacemaker Found in Back-Up Mode	30/08/2019	
					CT2985	Active	F1905	Device Revision or Replacement	A071205	Pacemaker Found in Back-Up Mode	30/08/2019	
					CT3001	Implantable metal	F1905	Device Revision or Replacement	A071205	Pacemaker Found in Back-Up Mode	30/08/2019	
					CT3002	Implantable synthetic polymer	F1905	Device Revision or Replacement	A071205	Pacemaker Found in Back-Up Mode	30/08/2019	
					CT3006	Implantable inorganic material	F1905	Device Revision or Replacement	A071205	Pacemaker Found in Back-Up Mode	30/08/2019	
					CT314	Electrically powered	F1905	Device Revision or Replacement	A071205	Pacemaker Found in Back-Up Mode	30/08/2019	
					CT321	Long-term surgical invasive	F1905	Device Revision or Replacement	A071205	Pacemaker Found in Back-Up Mode	30/08/2019	
					CT449	Generators and associated devices	F1905	Device Revision or Replacement	A071205	Pacemaker Found in Back-Up Mode	30/08/2019	
					CT981	Single use	F1905	Device Revision or Replacement	A071205	Pacemaker Found in Back-Up Mode	30/08/2019	
					CT983	Surgical invasive	F1905	Device Revision or Replacement	A071205	Pacemaker Found in Back-Up Mode	30/08/2019	
9751	2020	All other reportable incidents	17882	Physiologic-monitoring defibrillation system	CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A07	Electrical /Electronic Property Problem		
							F27	Problem Identified During Non-Clinical Procedure	A07	Electrical /Electronic Property Problem		
9752	2019	All other reportable incidents	35852	Single-chamber implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A072201	High Impedance		
9753	2019	All other reportable incidents	17882	Physiologic-monitoring defibrillation system	CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A071301	Failure to Deliver Shock/Stimulation		
							F10	Inadequate/inappropriate Treatment or Diagnostic Exposure	A071301	Failure to Deliver Shock/Stimulation		
							F26	No Health Consequences or Impact	A071301	Failure to Deliver Shock/Stimulation		
9754	2019	All other reportable incidents	36241	Pacemaker/defibrillator lead	CT1684	Cardiac pacemakers and associated devices	F1905	Device Revision or Replacement	A27	Appropriate Device Problem Term/Code not Available		
9755	2019	All other reportable incidents	36241	Pacemaker/defibrillator lead	CT1684	Cardiac pacemakers and associated devices	E2114	Perforation	A070102	High Capture Threshold		
							F1905	Device Revision or Replacement	A070102	High Capture Threshold		
9756	2019	All other reportable incidents	47266	Single-chamber implantable pacemaker, fixed-rate/demand	CT141	Electrodes/Leads and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070909	Oversensing		
					CT145	Cardiovascular devices	E2403	No Clinical Signs, Symptoms or Conditions	A070909	Oversensing		
					CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070909	Oversensing		
					CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A070909	Oversensing		
					CT233	Surgical	E2403	No Clinical Signs, Symptoms or Conditions	A070909	Oversensing		
					CT2406	Implantable	E2403	No Clinical Signs, Symptoms or Conditions	A070909	Oversensing		
					CT2407	Active implantable	E2403	No Clinical Signs, Symptoms or Conditions	A070909	Oversensing		
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A070909	Oversensing		
					CT2727	Active implant devices/ancillaries	E2403	No Clinical Signs, Symptoms or Conditions	A070909	Oversensing		
					CT2750	Implantable cardiac electrical stimulators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070909	Oversensing		
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A070909	Oversensing		
					CT2985	Active	E2403	No Clinical Signs, Symptoms or Conditions	A070909	Oversensing		
					CT3001	Implantable metal	E2403	No Clinical Signs, Symptoms or Conditions	A070909	Oversensing		
					CT3002	Implantable synthetic polymer	E2403	No Clinical Signs, Symptoms or Conditions	A070909	Oversensing		
					CT3006	Implantable inorganic material	E2403	No Clinical Signs, Symptoms or Conditions	A070909	Oversensing		
					CT314	Electrically powered	E2403	No Clinical Signs, Symptoms or Conditions	A070909	Oversensing		
					CT321	Long-term surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A070909	Oversensing		
					CT449	Generators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070909	Oversensing		
					CT981	Single use	E2403	No Clinical Signs, Symptoms or Conditions	A070909	Oversensing		
					CT983	Surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A070909	Oversensing		
9757	2020	All other reportable incidents	35852	Single-chamber implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	11/08/2020	11/08/2020
							F1901	Additional Surgery	A070504	Premature Discharge of Battery	11/08/2020	11/08/2020
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	11/08/2020	11/08/2020

9758	2019	All other reportable incidents	37265	Dual-chamber implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery		
9759	2019	All other reportable incidents	35852	Single-chamber implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A070909	Oversensing		
9760	2019	Death	36241	Pacemaker/defibrillator lead	CT1684	Cardiac pacemakers and associated devices	F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A090801	Signal Artifact/Noise	06/09/2019	06/09/2019
9761	2020	All other reportable incidents	35852	Single-chamber implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	07/10/2020	07/10/2020
									A071302	Inappropriate/Inadequate Shock/Stimulation	07/10/2020	07/10/2020
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	07/10/2020	07/10/2020
									A071302	Inappropriate/Inadequate Shock/Stimulation	07/10/2020	07/10/2020
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	07/10/2020	07/10/2020
									A071302	Inappropriate/Inadequate Shock/Stimulation	07/10/2020	07/10/2020
9762	2020	All other reportable incidents	36241	Pacemaker/defibrillator lead	CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070909	Oversensing		
									A090801	Signal Artifact/Noise		
									A1301	Failure to Read Input Signal		
							F19	Surgical Intervention	A070909	Oversensing		
									A090801	Signal Artifact/Noise		
									A1301	Failure to Read Input Signal		
9763	2020	All other reportable incidents	35852	Single-chamber implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A072201	High Impedance		
							F24	Insufficient Health Impact Information	A072201	High Impedance		
9764	2020	All other reportable incidents	36241	Pacemaker/defibrillator lead	CT1684	Cardiac pacemakers and associated devices	E0112	Dizziness	A05	Mechanical Problem	05/08/2020	05/08/2020
									A051201	Device Dislodged or Dislocated	05/08/2020	05/08/2020
									A070102	High Capture Threshold	05/08/2020	05/08/2020
									A2304	Off-Label Use	05/08/2020	05/08/2020
							E060104	Bradycardia	A05	Mechanical Problem	05/08/2020	05/08/2020
									A051201	Device Dislodged or Dislocated	05/08/2020	05/08/2020
									A070102	High Capture Threshold	05/08/2020	05/08/2020
									A2304	Off-Label Use	05/08/2020	05/08/2020
							F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A05	Mechanical Problem	05/08/2020	05/08/2020
									A051201	Device Dislodged or Dislocated	05/08/2020	05/08/2020
									A070102	High Capture Threshold	05/08/2020	05/08/2020
									A2304	Off-Label Use	05/08/2020	05/08/2020
							F19	Surgical Intervention	A05	Mechanical Problem	05/08/2020	05/08/2020
									A051201	Device Dislodged or Dislocated	05/08/2020	05/08/2020
									A070102	High Capture Threshold	05/08/2020	05/08/2020
									A2304	Off-Label Use	05/08/2020	05/08/2020
9765	2019	All other reportable incidents	47270	Cardiac resynchronization therapy implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A072201	High Impedance		
9766	2019	All other reportable incidents	36241	Pacemaker/defibrillator lead	CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A072201	High Impedance		
9767	2019	All other reportable incidents	35852	Single-chamber implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A071302	Inappropriate/Inadequate Shock/Stimulation		
9768	2020	All other reportable incidents	47265	Dual-chamber implantable pacemaker, rate-responsive	CT141	Electrodes/Leads and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	21/09/2020	21/09/2020
							F19	Surgical Intervention	A070504	Premature Discharge of Battery	21/09/2020	21/09/2020
					CT145	Cardiovascular devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	21/09/2020	21/09/2020
							F19	Surgical Intervention	A070504	Premature Discharge of Battery	21/09/2020	21/09/2020
					CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	21/09/2020	21/09/2020
							F19	Surgical Intervention	A070504	Premature Discharge of Battery	21/09/2020	21/09/2020
					CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	21/09/2020	21/09/2020
							F19	Surgical Intervention	A070504	Premature Discharge of Battery	21/09/2020	21/09/2020
					CT233	Surgical	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	21/09/2020	21/09/2020
							F19	Surgical Intervention	A070504	Premature Discharge of Battery	21/09/2020	21/09/2020
					CT2406	Implantable	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	21/09/2020	21/09/2020
							F19	Surgical Intervention	A070504	Premature Discharge of Battery	21/09/2020	21/09/2020
					CT2407	Active implantable	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	21/09/2020	21/09/2020
							F19	Surgical Intervention	A070504	Premature Discharge of Battery	21/09/2020	21/09/2020
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	21/09/2020	21/09/2020
							F19	Surgical Intervention	A070504	Premature Discharge of Battery	21/09/2020	21/09/2020
					CT2727	Active implant devices/ancillaries	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	21/09/2020	21/09/2020
							F19	Surgical Intervention	A070504	Premature Discharge of Battery	21/09/2020	21/09/2020
					CT2750	Implantable cardiac electrical stimulators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	21/09/2020	21/09/2020
							F19	Surgical Intervention	A070504	Premature Discharge of Battery	21/09/2020	21/09/2020
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	21/09/2020	21/09/2020
							F19	Surgical Intervention	A070504	Premature Discharge of Battery	21/09/2020	21/09/2020
					CT2985	Active	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	21/09/2020	21/09/2020
							F19	Surgical Intervention	A070504	Premature Discharge of Battery	21/09/2020	21/09/2020
					CT3001	Implantable metal	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	21/09/2020	21/09/2020
							F19	Surgical Intervention	A070504	Premature Discharge of Battery	21/09/2020	21/09/2020
					CT3002	Implantable synthetic polymer	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	21/09/2020	21/09/2020
							F19	Surgical Intervention	A070504	Premature Discharge of Battery	21/09/2020	21/09/2020
					CT3006	Implantable inorganic material	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	21/09/2020	21/09/2020
							F19	Surgical Intervention	A070504	Premature Discharge of Battery	21/09/2020	21/09/2020
					CT314	Electrically powered	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	21/09/2020	21/09/2020
							F19	Surgical Intervention	A070504	Premature Discharge of Battery	21/09/2020	21/09/2020
					CT321	Long-term surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	21/09/2020	21/09/2020
							F19	Surgical Intervention	A070504	Premature Discharge of Battery	21/09/2020	21/09/2020
					CT449	Generators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	21/09/2020	21/09/2020
							F19	Surgical Intervention	A070504	Premature Discharge of Battery	21/09/2020	21/09/2020
					CT981	Single use	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	21/09/2020	21/09/2020
							F19	Surgical Intervention	A070504	Premature Discharge of Battery	21/09/2020	21/09/2020
					CT983	Surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	21/09/2020	21/09/2020
							F19	Surgical Intervention	A070504	Premature Discharge of Battery	21/09/2020	21/09/2020
9769	2020	All other reportable incidents	47270	Cardiac resynchronization therapy implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E2401	Insufficient Information	A072201	High Impedance		
							E2403	No Clinical Signs, Symptoms or Conditions	A072201	High Impedance		
							F26	No Health Consequences or Impact	A072201	High Impedance		
9770	2020	All other reportable incidents	36241	Pacemaker/defibrillator lead	CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A051201	Device Dislodged or Dislocated		
									A070101	Failure to Capture		
									A070910	Undersensing		
									A072202	Low Impedance		
							F1905	Device Revision or Replacement	A051201	Device Dislodged or Dislocated		
									A070101	Failure to Capture		
									A070910	Undersensing		
									A072202	Low Impedance		

							F26	No Health Consequences or Impact	A051201	Device Dislodged or Dislocated		
									A070101	Failure to Capture		
									A070910	Undersensing		
									A072202	Low Impedance		
9771	2020	All other reportable incidents	35852	Single-chamber implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	19/12/2020	19/12/2020
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	19/12/2020	19/12/2020
9772	2019	All other reportable incidents	47265	Dual-chamber implantable pacemaker, rate-responsive	CT141	Electrodes/Leads and associated devices	E0112	Dizziness	A	Annex A		
							E0602	Cardiac Arrest	A	Annex A		
							F1905	Device Revision or Replacement	A	Annex A		
					CT145	Cardiovascular devices	E0112	Dizziness	A	Annex A		
							E0602	Cardiac Arrest	A	Annex A		
							F1905	Device Revision or Replacement	A	Annex A		
					CT1684	Cardiac pacemakers and associated devices	E0112	Dizziness	A	Annex A		
							E0602	Cardiac Arrest	A	Annex A		
							F1905	Device Revision or Replacement	A	Annex A		
					CT1685	Cardiac pacemakers	E0112	Dizziness	A	Annex A		
							E0602	Cardiac Arrest	A	Annex A		
							F1905	Device Revision or Replacement	A	Annex A		
					CT233	Surgical	E0112	Dizziness	A	Annex A		
							E0602	Cardiac Arrest	A	Annex A		
							F1905	Device Revision or Replacement	A	Annex A		
					CT2406	Implantable	E0112	Dizziness	A	Annex A		
							E0602	Cardiac Arrest	A	Annex A		
							F1905	Device Revision or Replacement	A	Annex A		
					CT2407	Active implantable	E0112	Dizziness	A	Annex A		
							E0602	Cardiac Arrest	A	Annex A		
							F1905	Device Revision or Replacement	A	Annex A		
					CT257	Implantable cardiac pacemakers	E0112	Dizziness	A	Annex A		
							E0602	Cardiac Arrest	A	Annex A		
							F1905	Device Revision or Replacement	A	Annex A		
					CT2727	Active implant devices/ancillaries	E0112	Dizziness	A	Annex A		
							E0602	Cardiac Arrest	A	Annex A		
							F1905	Device Revision or Replacement	A	Annex A		
					CT2750	Implantable cardiac electrical stimulators and associated devices	E0112	Dizziness	A	Annex A		
							E0602	Cardiac Arrest	A	Annex A		
							F1905	Device Revision or Replacement	A	Annex A		
					CT2866	Implantable cardiac electrical stimulators	E0112	Dizziness	A	Annex A		
							E0602	Cardiac Arrest	A	Annex A		
							F1905	Device Revision or Replacement	A	Annex A		
					CT2985	Active	E0112	Dizziness	A	Annex A		
							E0602	Cardiac Arrest	A	Annex A		
							F1905	Device Revision or Replacement	A	Annex A		
					CT3001	Implantable metal	E0112	Dizziness	A	Annex A		
							E0602	Cardiac Arrest	A	Annex A		
							F1905	Device Revision or Replacement	A	Annex A		
					CT3002	Implantable synthetic polymer	E0112	Dizziness	A	Annex A		
							E0602	Cardiac Arrest	A	Annex A		
							F1905	Device Revision or Replacement	A	Annex A		
					CT3006	Implantable inorganic material	E0112	Dizziness	A	Annex A		
							E0602	Cardiac Arrest	A	Annex A		
							F1905	Device Revision or Replacement	A	Annex A		
					CT314	Electrically powered	E0112	Dizziness	A	Annex A		
							E0602	Cardiac Arrest	A	Annex A		
							F1905	Device Revision or Replacement	A	Annex A		
					CT321	Long-term surgical invasive	E0112	Dizziness	A	Annex A		
							E0602	Cardiac Arrest	A	Annex A		
							F1905	Device Revision or Replacement	A	Annex A		
					CT449	Generators and associated devices	E0112	Dizziness	A	Annex A		
							E0602	Cardiac Arrest	A	Annex A		
							F1905	Device Revision or Replacement	A	Annex A		
					CT981	Single use	E0112	Dizziness	A	Annex A		
							E0602	Cardiac Arrest	A	Annex A		
							F1905	Device Revision or Replacement	A	Annex A		
					CT983	Surgical invasive	E0112	Dizziness	A	Annex A		
							E0602	Cardiac Arrest	A	Annex A		
							F1905	Device Revision or Replacement	A	Annex A		
9773	2020	All other reportable incidents	17882	Physiologic-monitoring defibrillation system	CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070803	Failure to Power Up		
							F27	Problem Identified During Non-Clinical Procedure	A070803	Failure to Power Up		
9774	2019	All other reportable incidents	35223	Endocardial/interventricular septal pacing lead	CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A072202	Low Impedance		
9775	2019	All other reportable incidents	35852	Single-chamber implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	09/09/2019	
9776	2019	All other reportable incidents	35852	Single-chamber implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A071302	Inappropriate/Inadequate Shock/Stimulation		
9777	2019	All other reportable incidents	35852	Single-chamber implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A071302	Inappropriate/Inadequate Shock/Stimulation		
9778	2019	All other reportable incidents	36241	Pacemaker/defibrillator lead	CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A071302	Inappropriate/Inadequate Shock/Stimulation		
9779	2019	All other reportable incidents	47265	Dual-chamber implantable pacemaker, rate-responsive	CT141	Electrodes/Leads and associated devices	F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	20/09/2019	20/09/2019
					CT145	Cardiovascular devices	F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	20/09/2019	20/09/2019
					CT1684	Cardiac pacemakers and associated devices	F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	20/09/2019	20/09/2019
					CT1685	Cardiac pacemakers	F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	20/09/2019	20/09/2019
					CT233	Surgical	F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	20/09/2019	20/09/2019
					CT2406	Implantable	F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	20/09/2019	20/09/2019
					CT2407	Active implantable	F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	20/09/2019	20/09/2019
					CT257	Implantable cardiac pacemakers	F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	20/09/2019	20/09/2019
					CT2727	Active implant devices/ancillaries	F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	20/09/2019	20/09/2019
					CT2750	Implantable cardiac electrical stimulators and associated devices	F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	20/09/2019	20/09/2019
					CT2866	Implantable cardiac electrical stimulators	F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	20/09/2019	20/09/2019
					CT2985	Active	F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	20/09/2019	20/09/2019
					CT3001	Implantable metal	F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	20/09/2019	20/09/2019

					CT3002	Implantable synthetic polymer	F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	20/09/2019	20/09/2019
					CT3006	Implantable inorganic material	F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	20/09/2019	20/09/2019
					CT314	Electrically powered	F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	20/09/2019	20/09/2019
					CT321	Long-term surgical invasive	F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	20/09/2019	20/09/2019
					CT449	Generators and associated devices	F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	20/09/2019	20/09/2019
					CT981	Single use	F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	20/09/2019	20/09/2019
					CT983	Surgical invasive	F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	20/09/2019	20/09/2019
9780	2020	All other reportable incidents	35852	Single-chamber implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E2104	Electric Shock	A071302	Inappropriate/Inadequate Shock/Stimulation		
							E2403	No Clinical Signs, Symptoms or Conditions	A071302	Inappropriate/Inadequate Shock/Stimulation		
							F26	No Health Consequences or Impact	A071302	Inappropriate/Inadequate Shock/Stimulation		
9781	2020	All other reportable incidents	35852	Single-chamber implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E2104	Electric Shock	A071302	Inappropriate/Inadequate Shock/Stimulation		
							F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A071302	Inappropriate/Inadequate Shock/Stimulation		
							F1901	Additional Surgery	A071302	Inappropriate/Inadequate Shock/Stimulation		
							F1905	Device Revision or Replacement	A071302	Inappropriate/Inadequate Shock/Stimulation		
9782	2020	All other reportable incidents	36241	Pacemaker/defibrillator lead	CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070909	Oversensing		
									A0712	Pacing Problem		
							F26	No Health Consequences or Impact	A070909	Oversensing		
									A0712	Pacing Problem		
9783	2019	All other reportable incidents	35852	Single-chamber implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E060109	Tachycardia	A0712	Pacing Problem	17/02/2021	17/02/2021
									A071202	Inaccurate Synchronization	17/02/2021	17/02/2021
									A071302	Inappropriate/Inadequate Shock/Stimulation	17/02/2021	17/02/2021
							E060110	Ventricular Fibrillation	A0712	Pacing Problem	17/02/2021	17/02/2021
									A071202	Inaccurate Synchronization	17/02/2021	17/02/2021
									A071302	Inappropriate/Inadequate Shock/Stimulation	17/02/2021	17/02/2021
							E2403	No Clinical Signs, Symptoms or Conditions	A0712	Pacing Problem	17/02/2021	17/02/2021
									A071202	Inaccurate Synchronization	17/02/2021	17/02/2021
									A071302	Inappropriate/Inadequate Shock/Stimulation	17/02/2021	17/02/2021
							F26	No Health Consequences or Impact	A0712	Pacing Problem	17/02/2021	17/02/2021
									A071202	Inaccurate Synchronization	17/02/2021	17/02/2021
									A071302	Inappropriate/Inadequate Shock/Stimulation	17/02/2021	17/02/2021
9784	2019	All other reportable incidents	47265	Dual-chamber implantable pacemaker, rate-responsive	CT141	Electrodes/Leads and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A27	Appropriate Device Problem Term/Code not Available		
					CT145	Cardiovascular devices	E2403	No Clinical Signs, Symptoms or Conditions	A27	Appropriate Device Problem Term/Code not Available		
					CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A27	Appropriate Device Problem Term/Code not Available		
					CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A27	Appropriate Device Problem Term/Code not Available		
					CT233	Surgical	E2403	No Clinical Signs, Symptoms or Conditions	A27	Appropriate Device Problem Term/Code not Available		
					CT2406	Implantable	E2403	No Clinical Signs, Symptoms or Conditions	A27	Appropriate Device Problem Term/Code not Available		
					CT2407	Active implantable	E2403	No Clinical Signs, Symptoms or Conditions	A27	Appropriate Device Problem Term/Code not Available		
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A27	Appropriate Device Problem Term/Code not Available		
					CT2727	Active implant devices/ancillaries	E2403	No Clinical Signs, Symptoms or Conditions	A27	Appropriate Device Problem Term/Code not Available		
					CT2750	Implantable cardiac electrical stimulators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A27	Appropriate Device Problem Term/Code not Available		
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A27	Appropriate Device Problem Term/Code not Available		
					CT2985	Active	E2403	No Clinical Signs, Symptoms or Conditions	A27	Appropriate Device Problem Term/Code not Available		
					CT3001	Implantable metal	E2403	No Clinical Signs, Symptoms or Conditions	A27	Appropriate Device Problem Term/Code not Available		
					CT3002	Implantable synthetic polymer	E2403	No Clinical Signs, Symptoms or Conditions	A27	Appropriate Device Problem Term/Code not Available		
					CT3006	Implantable inorganic material	E2403	No Clinical Signs, Symptoms or Conditions	A27	Appropriate Device Problem Term/Code not Available		
					CT314	Electrically powered	E2403	No Clinical Signs, Symptoms or Conditions	A27	Appropriate Device Problem Term/Code not Available		
					CT321	Long-term surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A27	Appropriate Device Problem Term/Code not Available		
					CT449	Generators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A27	Appropriate Device Problem Term/Code not Available		
					CT981	Single use	E2403	No Clinical Signs, Symptoms or Conditions	A27	Appropriate Device Problem Term/Code not Available		
					CT983	Surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A27	Appropriate Device Problem Term/Code not Available		
9785	2019	Death	47270	Cardiac resynchronization therapy implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	F21	Unexpected Deterioration	A071102	Failure to Interrogate	26/09/2019	
9786	2020	All other reportable incidents	35852	Single-chamber implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A072201	High Impedance		
							F26	No Health Consequences or Impact	A072201	High Impedance		
9787	2019	All other reportable incidents	47270	Cardiac resynchronization therapy implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	13/09/2019	
9788	2019	All other reportable incidents	47265	Dual-chamber implantable pacemaker, rate-responsive	CT141	Electrodes/Leads and associated devices	F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery		
					CT145	Cardiovascular devices	F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery		
					CT1684	Cardiac pacemakers and associated devices	F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery		
					CT1685	Cardiac pacemakers	F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery		
					CT233	Surgical	F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery		
					CT2406	Implantable	F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery		
					CT2407	Active implantable	F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery		
					CT257	Implantable cardiac pacemakers	F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery		
					CT2727	Active implant devices/ancillaries	F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery		
					CT2750	Implantable cardiac electrical stimulators and associated devices	F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery		
					CT2866	Implantable cardiac electrical stimulators	F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery		
					CT2985	Active	F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery		
					CT3001	Implantable metal	F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery		
					CT3002	Implantable synthetic polymer	F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery		
					CT3006	Implantable inorganic material	F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery		
					CT314	Electrically powered	F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery		

					CT321	Long-term surgical invasive	F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery		
					CT449	Generators and associated devices	F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery		
					CT981	Single use	F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery		
					CT983	Surgical invasive	F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery		
9789	2019	All other reportable incidents	35852	Single-chamber implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A070909	Oversensing		
9790	2019	All other reportable incidents	47265	Dual-chamber implantable pacemaker, rate-responsive	CT141	Electrodes/Leads and associated devices	F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery		
					CT145	Cardiovascular devices	F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery		
					CT1684	Cardiac pacemakers and associated devices	F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery		
					CT1685	Cardiac pacemakers	F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery		
					CT233	Surgical	F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery		
					CT2406	Implantable	F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery		
					CT2407	Active implantable	F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery		
					CT257	Implantable cardiac pacemakers	F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery		
					CT2727	Active implant devices/ancillaries	F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery		
					CT2750	Implantable cardiac electrical stimulators and associated devices	F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery		
					CT2866	Implantable cardiac electrical stimulators	F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery		
					CT2985	Active	F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery		
					CT3001	Implantable metal	F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery		
					CT3002	Implantable synthetic polymer	F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery		
					CT3006	Implantable inorganic material	F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery		
					CT314	Electrically powered	F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery		
					CT321	Long-term surgical invasive	F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery		
					CT449	Generators and associated devices	F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery		
					CT981	Single use	F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery		
					CT983	Surgical invasive	F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery		
9791	2019	All other reportable incidents	36241	Pacemaker/defibrillator lead	CT1684	Cardiac pacemakers and associated devices	E2114	Perforation	A24	No Apparent Device Problem		
9792	2020	All other reportable incidents	35223	Endocardial/interventricular septal pacing lead	CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A040101	Fracture		
									A070909	Oversensing		
									A090801	Signal Artifact/Noise		
							F26	No Health Consequences or Impact	A040101	Fracture		
									A070909	Oversensing		
									A090801	Signal Artifact/Noise		
9793	2020	All other reportable incidents	47265	Dual-chamber implantable pacemaker, rate-responsive	CT141	Electrodes/Leads and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT145	Cardiovascular devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT233	Surgical	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT2406	Implantable	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT2407	Active implantable	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT2727	Active implant devices/ancillaries	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT2750	Implantable cardiac electrical stimulators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT2985	Active	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT3001	Implantable metal	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT3002	Implantable synthetic polymer	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT3006	Implantable inorganic material	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT314	Electrically powered	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT321	Long-term surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT449	Generators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT981	Single use	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		

						F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery			
						F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery			
				CT983	Surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery			
						F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery			
						F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery			
9794	2020	All other reportable incidents	47265	Dual-chamber implantable pacemaker, rate-responsive	CT141	Electrodes/Leads and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	13/08/2020	13/08/2020
							F12	Serious Injury/ Illness/ Impairment	A070504	Premature Discharge of Battery	13/08/2020	13/08/2020
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	13/08/2020	13/08/2020
					CT145	Cardiovascular devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	13/08/2020	13/08/2020
							F12	Serious Injury/ Illness/ Impairment	A070504	Premature Discharge of Battery	13/08/2020	13/08/2020
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	13/08/2020	13/08/2020
					CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	13/08/2020	13/08/2020
							F12	Serious Injury/ Illness/ Impairment	A070504	Premature Discharge of Battery	13/08/2020	13/08/2020
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	13/08/2020	13/08/2020
					CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	13/08/2020	13/08/2020
							F12	Serious Injury/ Illness/ Impairment	A070504	Premature Discharge of Battery	13/08/2020	13/08/2020
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	13/08/2020	13/08/2020
					CT233	Surgical	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	13/08/2020	13/08/2020
							F12	Serious Injury/ Illness/ Impairment	A070504	Premature Discharge of Battery	13/08/2020	13/08/2020
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	13/08/2020	13/08/2020
					CT2406	Implantable	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	13/08/2020	13/08/2020
							F12	Serious Injury/ Illness/ Impairment	A070504	Premature Discharge of Battery	13/08/2020	13/08/2020
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	13/08/2020	13/08/2020
					CT2407	Active implantable	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	13/08/2020	13/08/2020
							F12	Serious Injury/ Illness/ Impairment	A070504	Premature Discharge of Battery	13/08/2020	13/08/2020
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	13/08/2020	13/08/2020
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	13/08/2020	13/08/2020
							F12	Serious Injury/ Illness/ Impairment	A070504	Premature Discharge of Battery	13/08/2020	13/08/2020
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	13/08/2020	13/08/2020
					CT2727	Active implant devices/ancillaries	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	13/08/2020	13/08/2020
							F12	Serious Injury/ Illness/ Impairment	A070504	Premature Discharge of Battery	13/08/2020	13/08/2020
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	13/08/2020	13/08/2020
					CT2750	Implantable cardiac electrical stimulators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	13/08/2020	13/08/2020
							F12	Serious Injury/ Illness/ Impairment	A070504	Premature Discharge of Battery	13/08/2020	13/08/2020
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	13/08/2020	13/08/2020
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	13/08/2020	13/08/2020
							F12	Serious Injury/ Illness/ Impairment	A070504	Premature Discharge of Battery	13/08/2020	13/08/2020
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	13/08/2020	13/08/2020
					CT2985	Active	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	13/08/2020	13/08/2020
							F12	Serious Injury/ Illness/ Impairment	A070504	Premature Discharge of Battery	13/08/2020	13/08/2020
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	13/08/2020	13/08/2020
					CT3001	Implantable metal	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	13/08/2020	13/08/2020
							F12	Serious Injury/ Illness/ Impairment	A070504	Premature Discharge of Battery	13/08/2020	13/08/2020
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	13/08/2020	13/08/2020
					CT3002	Implantable synthetic polymer	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	13/08/2020	13/08/2020
							F12	Serious Injury/ Illness/ Impairment	A070504	Premature Discharge of Battery	13/08/2020	13/08/2020
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	13/08/2020	13/08/2020
					CT3006	Implantable inorganic material	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	13/08/2020	13/08/2020
							F12	Serious Injury/ Illness/ Impairment	A070504	Premature Discharge of Battery	13/08/2020	13/08/2020
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	13/08/2020	13/08/2020
					CT314	Electrically powered	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	13/08/2020	13/08/2020
							F12	Serious Injury/ Illness/ Impairment	A070504	Premature Discharge of Battery	13/08/2020	13/08/2020
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	13/08/2020	13/08/2020
					CT321	Long-term surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	13/08/2020	13/08/2020
							F12	Serious Injury/ Illness/ Impairment	A070504	Premature Discharge of Battery	13/08/2020	13/08/2020
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	13/08/2020	13/08/2020
					CT449	Generators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	13/08/2020	13/08/2020
							F12	Serious Injury/ Illness/ Impairment	A070504	Premature Discharge of Battery	13/08/2020	13/08/2020
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	13/08/2020	13/08/2020
					CT981	Single use	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	13/08/2020	13/08/2020
							F12	Serious Injury/ Illness/ Impairment	A070504	Premature Discharge of Battery	13/08/2020	13/08/2020
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	13/08/2020	13/08/2020
					CT983	Surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	13/08/2020	13/08/2020
							F12	Serious Injury/ Illness/ Impairment	A070504	Premature Discharge of Battery	13/08/2020	13/08/2020
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	13/08/2020	13/08/2020
9795	2020	All other reportable incidents	17882	Physiologic-monitoring defibrillation system	CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070912	Incorrect Interpretation of Signal		
							F26	No Health Consequences or Impact	A070912	Incorrect Interpretation of Signal		
9796	2020	All other reportable incidents	35852	Single-chamber implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E2104	Electric Shock	A071302	Inappropriate/Inadequate Shock/Stimulation	21/08/2020	21/08/2020
							F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A071302	Inappropriate/Inadequate Shock/Stimulation	21/08/2020	21/08/2020
							F1905	Device Revision or Replacement	A071302	Inappropriate/Inadequate Shock/Stimulation	21/08/2020	21/08/2020
9797	2020	All other reportable incidents	36241	Pacemaker/defibrillator lead	CT1684	Cardiac pacemakers and associated devices	E0506	Hemorrhage/Blood Loss/Bleeding	A051201	Device Dislodged or Dislocated		
									A070101	Failure to Capture		
							F12	Serious Injury/ Illness/ Impairment	A051201	Device Dislodged or Dislocated		
									A070101	Failure to Capture		
							F1905	Device Revision or Replacement	A051201	Device Dislodged or Dislocated		
									A070101	Failure to Capture		
9798	2020	All other reportable incidents	35852	Single-chamber implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E0138	Undesired Nerve Stimulation	A070909	Oversensing		
									A0712	Pacing Problem		
									A071209	Extracardiac Stimulation		
									A072201	High Impedance		
							F08	Hospitalization or Prolonged Hospitalization	A070909	Oversensing		
									A0712	Pacing Problem		
									A071209	Extracardiac Stimulation		
									A072201	High Impedance		
							F14	Prolonged Episode of Care	A070909	Oversensing		
									A0712	Pacing Problem		
									A071209	Extracardiac Stimulation		
									A072201	High Impedance		

							F23	Unexpected Medical Intervention	A070909	Oversensing		
									A0712	Pacing Problem		
									A071209	Extracardiac Stimulation		
									A072201	High Impedance		
9799	2020	All other reportable incidents	47265	Dual-chamber implantable pacemaker, rate-responsive	CT141	Electrodes/Leads and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem		
							F26	No Health Consequences or Impact	A0705	Battery Problem		
					CT145	Cardiovascular devices	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem		
							F26	No Health Consequences or Impact	A0705	Battery Problem		
					CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem		
							F26	No Health Consequences or Impact	A0705	Battery Problem		
					CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem		
							F26	No Health Consequences or Impact	A0705	Battery Problem		
					CT233	Surgical	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem		
							F26	No Health Consequences or Impact	A0705	Battery Problem		
					CT2406	Implantable	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem		
							F26	No Health Consequences or Impact	A0705	Battery Problem		
					CT2407	Active implantable	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem		
							F26	No Health Consequences or Impact	A0705	Battery Problem		
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem		
							F26	No Health Consequences or Impact	A0705	Battery Problem		
					CT2727	Active implant devices/ancillaries	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem		
							F26	No Health Consequences or Impact	A0705	Battery Problem		
					CT2750	Implantable cardiac electrical stimulators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem		
							F26	No Health Consequences or Impact	A0705	Battery Problem		
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem		
							F26	No Health Consequences or Impact	A0705	Battery Problem		
					CT2985	Active	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem		
							F26	No Health Consequences or Impact	A0705	Battery Problem		
					CT3001	Implantable metal	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem		
							F26	No Health Consequences or Impact	A0705	Battery Problem		
					CT3002	Implantable synthetic polymer	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem		
							F26	No Health Consequences or Impact	A0705	Battery Problem		
					CT3006	Implantable inorganic material	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem		
							F26	No Health Consequences or Impact	A0705	Battery Problem		
					CT314	Electrically powered	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem		
							F26	No Health Consequences or Impact	A0705	Battery Problem		
					CT321	Long-term surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem		
							F26	No Health Consequences or Impact	A0705	Battery Problem		
					CT449	Generators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem		
							F26	No Health Consequences or Impact	A0705	Battery Problem		
					CT981	Single use	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem		
							F26	No Health Consequences or Impact	A0705	Battery Problem		
					CT983	Surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem		
							F26	No Health Consequences or Impact	A0705	Battery Problem		
9800	2020	All other reportable incidents	47270	Cardiac resynchronization therapy implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A070101	Failure to Capture		
									A070102	High Capture Threshold		
									A072201	High Impedance		
							F15	Recognized Device or Procedural Complication	A070101	Failure to Capture		
									A070102	High Capture Threshold		
									A072201	High Impedance		
9801	2020	All other reportable incidents	37265	Dual-chamber implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F19	Surgical Intervention	A070504	Premature Discharge of Battery		
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery		
9802	2020	All other reportable incidents	37265	Dual-chamber implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A071302	Inappropriate/Inadequate Shock/Stimulation		
							F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A071302	Inappropriate/Inadequate Shock/Stimulation		
							F23	Unexpected Medical Intervention	A071302	Inappropriate/Inadequate Shock/Stimulation		
9803	2020	All other reportable incidents	37265	Dual-chamber implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	14/08/2020	14/08/2020
							F1901	Additional Surgery	A070504	Premature Discharge of Battery	14/08/2020	14/08/2020
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	14/08/2020	14/08/2020
9804	2020	Death	47270	Cardiac resynchronization therapy implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E2401	Insufficient Information	A0709	Device Sensing Problem		
							F02	Death	A0709	Device Sensing Problem		
							F1203	Life Threatening Illness or Injury	A0709	Device Sensing Problem		
9805	2020	Death	35223	Endocardial/interventricular septal pacing lead	CT1684	Cardiac pacemakers and associated devices	E2401	Insufficient Information	A0722	Impedance Problem		
									A072201	High Impedance		
							F02	Death	A0722	Impedance Problem		
									A072201	High Impedance		
9806	2019	All other reportable incidents	36241	Pacemaker/defibrillator lead	CT1684	Cardiac pacemakers and associated devices	F1905	Device Revision or Replacement	A072201	High Impedance		
9807	2019	All other reportable incidents	35852	Single-chamber implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	F1905	Device Revision or Replacement	A010402	Migration		
									A070909	Oversensing		
									A071302	Inappropriate/Inadequate Shock/Stimulation		
9808	2020	All other reportable incidents	47265	Dual-chamber implantable pacemaker, rate-responsive	CT141	Electrodes/Leads and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT145	Cardiovascular devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT233	Surgical	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT2406	Implantable	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT2407	Active implantable	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT2727	Active implant devices/ancillaries	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		

					CT2750	Implantable cardiac electrical stimulators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT2985	Active	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT3001	Implantable metal	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT3002	Implantable synthetic polymer	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT3006	Implantable inorganic material	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT314	Electrically powered	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT321	Long-term surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT449	Generators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT981	Single use	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT983	Surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
9809	2020	All other reportable incidents	35223	Endocardial/interventricular septal pacing lead	CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A07	Electrical /Electronic Property Problem		
									A072202	Low Impedance		
							F26	No Health Consequences or Impact	A07	Electrical /Electronic Property Problem		
									A072202	Low Impedance		
9810	2019	All other reportable incidents	47265	Dual-chamber implantable pacemaker, rate-responsive	CT141	Electrodes/Leads and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
					CT145	Cardiovascular devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
					CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
					CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
					CT233	Surgical	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
					CT2406	Implantable	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
					CT2407	Active implantable	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
					CT2727	Active implant devices/ancillaries	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
					CT2750	Implantable cardiac electrical stimulators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
					CT2985	Active	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
					CT3001	Implantable metal	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
					CT3002	Implantable synthetic polymer	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
					CT3006	Implantable inorganic material	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
					CT314	Electrically powered	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
					CT321	Long-term surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
					CT449	Generators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
					CT981	Single use	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
					CT983	Surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
9811	2019	All other reportable incidents	47265	Dual-chamber implantable pacemaker, rate-responsive	CT141	Electrodes/Leads and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
					CT145	Cardiovascular devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
					CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
					CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
					CT233	Surgical	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
					CT2406	Implantable	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
					CT2407	Active implantable	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
					CT2727	Active implant devices/ancillaries	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
					CT2750	Implantable cardiac electrical stimulators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
					CT2985	Active	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
					CT3001	Implantable metal	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
					CT3002	Implantable synthetic polymer	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
					CT3006	Implantable inorganic material	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
					CT314	Electrically powered	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
					CT321	Long-term surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
					CT449	Generators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
					CT981	Single use	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
					CT983	Surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
9812	2019	All other reportable incidents	37265	Dual-chamber implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	17/02/2021	17/02/2021
							F19	Surgical Intervention	A070504	Premature Discharge of Battery	17/02/2021	17/02/2021
9813	2020	All other reportable incidents	17882	Physiologic-monitoring defibrillation system	CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070803	Failure to Power Up		
							F27	Problem Identified During Non-Clinical Procedure	A070803	Failure to Power Up		
9814	2019	All other reportable incidents	47270	Cardiac resynchronization therapy implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A072201	High Impedance		
9815	2019	All other reportable incidents	37265	Dual-chamber implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A070909	Oversensing		
9816	2019	All other reportable incidents	35852	Single-chamber implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A070909	Oversensing		
9817	2020	All other reportable incidents	47263	Cardiac resynchronization therapy implantable pacemaker	CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	12/08/2020	12/08/2020
							F1901	Additional Surgery	A070504	Premature Discharge of Battery	12/08/2020	12/08/2020
					CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	12/08/2020	12/08/2020
							F1901	Additional Surgery	A070504	Premature Discharge of Battery	12/08/2020	12/08/2020
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	12/08/2020	12/08/2020
							F1901	Additional Surgery	A070504	Premature Discharge of Battery	12/08/2020	12/08/2020
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	12/08/2020	12/08/2020
							F1901	Additional Surgery	A070504	Premature Discharge of Battery	12/08/2020	12/08/2020
9818	2020	All other reportable incidents	37265	Dual-chamber implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A070910	Undersensing		
									A0712	Pacing Problem		
							F26	No Health Consequences or Impact	A070910	Undersensing		
									A0712	Pacing Problem		

9819	2019	All other reportable incidents	37265	Dual-chamber implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
9820	2020	All other reportable incidents	47265	Dual-chamber implantable pacemaker, rate-responsive	CT141	Electrodes/Leads and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT145	Cardiovascular devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT233	Surgical	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT2406	Implantable	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT2407	Active implantable	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT2727	Active implant devices/ancillaries	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT2750	Implantable cardiac electrical stimulators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT2985	Active	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT3001	Implantable metal	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT3002	Implantable synthetic polymer	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT3006	Implantable inorganic material	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT314	Electrically powered	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT321	Long-term surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT449	Generators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT981	Single use	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT983	Surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
9821	2020	All other reportable incidents	47265	Dual-chamber implantable pacemaker, rate-responsive	CT141	Electrodes/Leads and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT145	Cardiovascular devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT233	Surgical	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT2406	Implantable	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT2407	Active implantable	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT2727	Active implant devices/ancillaries	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT2750	Implantable cardiac electrical stimulators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT2985	Active	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT3001	Implantable metal	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT3002	Implantable synthetic polymer	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT3006	Implantable inorganic material	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT314	Electrically powered	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT321	Long-term surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT449	Generators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT981	Single use	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT983	Surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
9822	2020	All other reportable incidents	36241	Pacemaker/defibrillator lead	CT1684	Cardiac pacemakers and associated devices	E0138	Undesired Nerve Stimulation	A0712	Pacing Problem		
									A071209	Extracardiac Stimulation		
									A072201	High Impedance		
									A090801	Signal Artifact/Noise		
							F05	Delay to Treatment/ Therapy	A0712	Pacing Problem		
									A071209	Extracardiac Stimulation		
									A072201	High Impedance		

									A090801	Signal Artifact/Noise		
							F08	Hospitalization or Prolonged Hospitalization	A0712	Pacing Problem		
									A071209	Extracardiac Stimulation		
									A072201	High Impedance		
									A090801	Signal Artifact/Noise		
							F23	Unexpected Medical Intervention	A0712	Pacing Problem		
									A071209	Extracardiac Stimulation		
									A072201	High Impedance		
									A090801	Signal Artifact/Noise		
9823	2019	All other reportable incidents	47267	Single-chamber implantable pacemaker, rate-responsive	CT141	Electrodes/Leads and associated devices	F1905	Device Revision or Replacement	A160601	Premature Elective Replacement Indicator		
					CT145	Cardiovascular devices	F1905	Device Revision or Replacement	A160601	Premature Elective Replacement Indicator		
					CT1684	Cardiac pacemakers and associated devices	F1905	Device Revision or Replacement	A160601	Premature Elective Replacement Indicator		
					CT1685	Cardiac pacemakers	F1905	Device Revision or Replacement	A160601	Premature Elective Replacement Indicator		
					CT233	Surgical	F1905	Device Revision or Replacement	A160601	Premature Elective Replacement Indicator		
					CT2406	Implantable	F1905	Device Revision or Replacement	A160601	Premature Elective Replacement Indicator		
					CT2407	Active implantable	F1905	Device Revision or Replacement	A160601	Premature Elective Replacement Indicator		
					CT257	Implantable cardiac pacemakers	F1905	Device Revision or Replacement	A160601	Premature Elective Replacement Indicator		
					CT2727	Active implant devices/ancillaries	F1905	Device Revision or Replacement	A160601	Premature Elective Replacement Indicator		
					CT2750	Implantable cardiac electrical stimulators and associated devices	F1905	Device Revision or Replacement	A160601	Premature Elective Replacement Indicator		
					CT2866	Implantable cardiac electrical stimulators	F1905	Device Revision or Replacement	A160601	Premature Elective Replacement Indicator		
					CT2985	Active	F1905	Device Revision or Replacement	A160601	Premature Elective Replacement Indicator		
					CT3001	Implantable metal	F1905	Device Revision or Replacement	A160601	Premature Elective Replacement Indicator		
					CT3002	Implantable synthetic polymer	F1905	Device Revision or Replacement	A160601	Premature Elective Replacement Indicator		
					CT3006	Implantable inorganic material	F1905	Device Revision or Replacement	A160601	Premature Elective Replacement Indicator		
					CT314	Electrically powered	F1905	Device Revision or Replacement	A160601	Premature Elective Replacement Indicator		
					CT321	Long-term surgical invasive	F1905	Device Revision or Replacement	A160601	Premature Elective Replacement Indicator		
					CT449	Generators and associated devices	F1905	Device Revision or Replacement	A160601	Premature Elective Replacement Indicator		
					CT981	Single use	F1905	Device Revision or Replacement	A160601	Premature Elective Replacement Indicator		
					CT983	Surgical invasive	F1905	Device Revision or Replacement	A160601	Premature Elective Replacement Indicator		
9824	2019	All other reportable incidents	37265	Dual-chamber implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	F1905	Device Revision or Replacement	A05	Mechanical Problem	01/07/2019	
9825	2019	All other reportable incidents	36241	Pacemaker/defibrillator lead	CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A	Annex A		
9826	2019	All other reportable incidents	36241	Pacemaker/defibrillator lead	CT1684	Cardiac pacemakers and associated devices	F1905	Device Revision or Replacement	A070910	Undersensing	05/09/2019	
9827	2019	Unanticipated serious deterioration in state of health	17882	Physiologic-monitoring defibrillation system	CT1684	Cardiac pacemakers and associated devices	F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A071302	Inappropriate/Inadequate Shock/Stimulation		
9828	2019	All other reportable incidents	47265	Dual-chamber implantable pacemaker, rate-responsive	CT141	Electrodes/Leads and associated devices	E0752	Unspecified Respiratory Problem	A070910	Undersensing		
					CT145	Cardiovascular devices	E0752	Unspecified Respiratory Problem	A070910	Undersensing		
					CT1684	Cardiac pacemakers and associated devices	E0752	Unspecified Respiratory Problem	A070910	Undersensing		
					CT1685	Cardiac pacemakers	E0752	Unspecified Respiratory Problem	A070910	Undersensing		
					CT233	Surgical	E0752	Unspecified Respiratory Problem	A070910	Undersensing		
					CT2406	Implantable	E0752	Unspecified Respiratory Problem	A070910	Undersensing		
					CT2407	Active implantable	E0752	Unspecified Respiratory Problem	A070910	Undersensing		
					CT257	Implantable cardiac pacemakers	E0752	Unspecified Respiratory Problem	A070910	Undersensing		
					CT2727	Active implant devices/ancillaries	E0752	Unspecified Respiratory Problem	A070910	Undersensing		
					CT2750	Implantable cardiac electrical stimulators and associated devices	E0752	Unspecified Respiratory Problem	A070910	Undersensing		
					CT2866	Implantable cardiac electrical stimulators	E0752	Unspecified Respiratory Problem	A070910	Undersensing		
					CT2985	Active	E0752	Unspecified Respiratory Problem	A070910	Undersensing		
					CT3001	Implantable metal	E0752	Unspecified Respiratory Problem	A070910	Undersensing		
					CT3002	Implantable synthetic polymer	E0752	Unspecified Respiratory Problem	A070910	Undersensing		
					CT3006	Implantable inorganic material	E0752	Unspecified Respiratory Problem	A070910	Undersensing		
					CT314	Electrically powered	E0752	Unspecified Respiratory Problem	A070910	Undersensing		
					CT321	Long-term surgical invasive	E0752	Unspecified Respiratory Problem	A070910	Undersensing		
					CT449	Generators and associated devices	E0752	Unspecified Respiratory Problem	A070910	Undersensing		
					CT981	Single use	E0752	Unspecified Respiratory Problem	A070910	Undersensing		
					CT983	Surgical invasive	E0752	Unspecified Respiratory Problem	A070910	Undersensing		
9829	2020	All other reportable incidents	47106	Drug-eluting endocardial pacing lead	CT1684	Cardiac pacemakers and associated devices	E0112	Dizziness	A0701	Capturing Problem		
									A070102	High Capture Threshold		
							E2007	Fall	A0701	Capturing Problem		
									A070102	High Capture Threshold		
							F1905	Device Revision or Replacement	A0701	Capturing Problem		
									A070102	High Capture Threshold		
9830	2019	All other reportable incidents	37265	Dual-chamber implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A071302	Inappropriate/Inadequate Shock/Stimulation		
9831	2019	All other reportable incidents	35852	Single-chamber implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A070910	Undersensing		
9832	2020	All other reportable incidents	36241	Pacemaker/defibrillator lead	CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070102	High Capture Threshold	12/11/2020	12/11/2020
									A070909	Oversensing	12/11/2020	12/11/2020
									A072201	High Impedance	12/11/2020	12/11/2020
									A1301	Failure to Read Input Signal	12/11/2020	12/11/2020
							F08	Hospitalization or Prolonged Hospitalization	A070102	High Capture Threshold	12/11/2020	12/11/2020
									A070909	Oversensing	12/11/2020	12/11/2020
									A072201	High Impedance	12/11/2020	12/11/2020
									A1301	Failure to Read Input Signal	12/11/2020	12/11/2020
							F15	Recognized Device or Procedural Complication	A070102	High Capture Threshold	12/11/2020	12/11/2020
									A070909	Oversensing	12/11/2020	12/11/2020
									A072201	High Impedance	12/11/2020	12/11/2020
									A1301	Failure to Read Input Signal	12/11/2020	12/11/2020
							F19	Surgical Intervention	A070102	High Capture Threshold	12/11/2020	12/11/2020
									A070909	Oversensing	12/11/2020	12/11/2020
									A072201	High Impedance	12/11/2020	12/11/2020
									A1301	Failure to Read Input Signal	12/11/2020	12/11/2020
9833	2020	All other reportable incidents	36241	Pacemaker/defibrillator lead	CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A072201	High Impedance	12/11/2020	12/11/2020
									A090801	Signal Artifact/Noise	12/11/2020	12/11/2020
									A1301	Failure to Read Input Signal	12/11/2020	12/11/2020
							F15	Recognized Device or Procedural Complication	A072201	High Impedance	12/11/2020	12/11/2020
									A090801	Signal Artifact/Noise	12/11/2020	12/11/2020
									A1301	Failure to Read Input Signal	12/11/2020	12/11/2020
							F19	Surgical Intervention	A072201	High Impedance	12/11/2020	12/11/2020
									A090801	Signal Artifact/Noise	12/11/2020	12/11/2020

									A1301	Failure to Read Input Signal	12/11/2020	12/11/2020
9834	2020	All other reportable incidents	47265	Dual-chamber implantable pacemaker, rate-responsive	CT141	Electrodes/Leads and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	02/12/2020	02/12/2020
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	02/12/2020	02/12/2020
							F12	Serious Injury/ Illness/ Impairment	A070504	Premature Discharge of Battery	02/12/2020	02/12/2020
							F1903	Device Explantation	A070504	Premature Discharge of Battery	02/12/2020	02/12/2020
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	02/12/2020	02/12/2020
					CT145	Cardiovascular devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	02/12/2020	02/12/2020
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	02/12/2020	02/12/2020
							F12	Serious Injury/ Illness/ Impairment	A070504	Premature Discharge of Battery	02/12/2020	02/12/2020
							F1903	Device Explantation	A070504	Premature Discharge of Battery	02/12/2020	02/12/2020
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	02/12/2020	02/12/2020
					CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	02/12/2020	02/12/2020
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	02/12/2020	02/12/2020
							F12	Serious Injury/ Illness/ Impairment	A070504	Premature Discharge of Battery	02/12/2020	02/12/2020
							F1903	Device Explantation	A070504	Premature Discharge of Battery	02/12/2020	02/12/2020
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	02/12/2020	02/12/2020
					CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	02/12/2020	02/12/2020
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	02/12/2020	02/12/2020
							F12	Serious Injury/ Illness/ Impairment	A070504	Premature Discharge of Battery	02/12/2020	02/12/2020
							F1903	Device Explantation	A070504	Premature Discharge of Battery	02/12/2020	02/12/2020
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	02/12/2020	02/12/2020
					CT233	Surgical	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	02/12/2020	02/12/2020
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	02/12/2020	02/12/2020
							F12	Serious Injury/ Illness/ Impairment	A070504	Premature Discharge of Battery	02/12/2020	02/12/2020
							F1903	Device Explantation	A070504	Premature Discharge of Battery	02/12/2020	02/12/2020
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	02/12/2020	02/12/2020
					CT2406	Implantable	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	02/12/2020	02/12/2020
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	02/12/2020	02/12/2020
							F12	Serious Injury/ Illness/ Impairment	A070504	Premature Discharge of Battery	02/12/2020	02/12/2020
							F1903	Device Explantation	A070504	Premature Discharge of Battery	02/12/2020	02/12/2020
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	02/12/2020	02/12/2020
					CT2407	Active implantable	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	02/12/2020	02/12/2020
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	02/12/2020	02/12/2020
							F12	Serious Injury/ Illness/ Impairment	A070504	Premature Discharge of Battery	02/12/2020	02/12/2020
							F1903	Device Explantation	A070504	Premature Discharge of Battery	02/12/2020	02/12/2020
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	02/12/2020	02/12/2020
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	02/12/2020	02/12/2020
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	02/12/2020	02/12/2020
							F12	Serious Injury/ Illness/ Impairment	A070504	Premature Discharge of Battery	02/12/2020	02/12/2020
							F1903	Device Explantation	A070504	Premature Discharge of Battery	02/12/2020	02/12/2020
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	02/12/2020	02/12/2020
					CT2727	Active implant devices/ancillaries	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	02/12/2020	02/12/2020
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	02/12/2020	02/12/2020
							F12	Serious Injury/ Illness/ Impairment	A070504	Premature Discharge of Battery	02/12/2020	02/12/2020
							F1903	Device Explantation	A070504	Premature Discharge of Battery	02/12/2020	02/12/2020
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	02/12/2020	02/12/2020
					CT2750	Implantable cardiac electrical stimulators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	02/12/2020	02/12/2020
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	02/12/2020	02/12/2020
							F12	Serious Injury/ Illness/ Impairment	A070504	Premature Discharge of Battery	02/12/2020	02/12/2020
							F1903	Device Explantation	A070504	Premature Discharge of Battery	02/12/2020	02/12/2020
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	02/12/2020	02/12/2020
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	02/12/2020	02/12/2020
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	02/12/2020	02/12/2020
							F12	Serious Injury/ Illness/ Impairment	A070504	Premature Discharge of Battery	02/12/2020	02/12/2020
							F1903	Device Explantation	A070504	Premature Discharge of Battery	02/12/2020	02/12/2020
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	02/12/2020	02/12/2020
					CT2985	Active	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	02/12/2020	02/12/2020
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	02/12/2020	02/12/2020
							F12	Serious Injury/ Illness/ Impairment	A070504	Premature Discharge of Battery	02/12/2020	02/12/2020
							F1903	Device Explantation	A070504	Premature Discharge of Battery	02/12/2020	02/12/2020
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	02/12/2020	02/12/2020
					CT3001	Implantable metal	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	02/12/2020	02/12/2020
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	02/12/2020	02/12/2020
							F12	Serious Injury/ Illness/ Impairment	A070504	Premature Discharge of Battery	02/12/2020	02/12/2020
							F1903	Device Explantation	A070504	Premature Discharge of Battery	02/12/2020	02/12/2020
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	02/12/2020	02/12/2020
					CT3002	Implantable synthetic polymer	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	02/12/2020	02/12/2020
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	02/12/2020	02/12/2020
							F12	Serious Injury/ Illness/ Impairment	A070504	Premature Discharge of Battery	02/12/2020	02/12/2020
							F1903	Device Explantation	A070504	Premature Discharge of Battery	02/12/2020	02/12/2020
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	02/12/2020	02/12/2020
					CT3006	Implantable inorganic material	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	02/12/2020	02/12/2020
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	02/12/2020	02/12/2020
							F12	Serious Injury/ Illness/ Impairment	A070504	Premature Discharge of Battery	02/12/2020	02/12/2020
							F1903	Device Explantation	A070504	Premature Discharge of Battery	02/12/2020	02/12/2020
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	02/12/2020	02/12/2020
					CT314	Electrically powered	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	02/12/2020	02/12/2020
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	02/12/2020	02/12/2020
							F12	Serious Injury/ Illness/ Impairment	A070504	Premature Discharge of Battery	02/12/2020	02/12/2020
							F1903	Device Explantation	A070504	Premature Discharge of Battery	02/12/2020	02/12/2020
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	02/12/2020	02/12/2020
					CT321	Long-term surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	02/12/2020	02/12/2020
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	02/12/2020	02/12/2020
							F12	Serious Injury/ Illness/ Impairment	A070504	Premature Discharge of Battery	02/12/2020	02/12/2020
							F1903	Device Explantation	A070504	Premature Discharge of Battery	02/12/2020	02/12/2020
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	02/12/2020	02/12/2020
					CT449	Generators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	02/12/2020	02/12/2020
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	02/12/2020	02/12/2020

							F12	Serious Injury/ Illness/ Impairment	A070504	Premature Discharge of Battery	02/12/2020	02/12/2020
							F1903	Device Explantation	A070504	Premature Discharge of Battery	02/12/2020	02/12/2020
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	02/12/2020	02/12/2020
					CT981	Single use	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	02/12/2020	02/12/2020
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	02/12/2020	02/12/2020
							F12	Serious Injury/ Illness/ Impairment	A070504	Premature Discharge of Battery	02/12/2020	02/12/2020
							F1903	Device Explantation	A070504	Premature Discharge of Battery	02/12/2020	02/12/2020
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	02/12/2020	02/12/2020
					CT983	Surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	02/12/2020	02/12/2020
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	02/12/2020	02/12/2020
							F12	Serious Injury/ Illness/ Impairment	A070504	Premature Discharge of Battery	02/12/2020	02/12/2020
							F1903	Device Explantation	A070504	Premature Discharge of Battery	02/12/2020	02/12/2020
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	02/12/2020	02/12/2020
9835	2019	All other reportable incidents	37265	Dual-chamber implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	F1905	Device Revision or Replacement	A0705	Battery Problem	20/09/2019	
9836	2020	All other reportable incidents	36241	Pacemaker/defibrillator lead	CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A072201	High Impedance	14/08/2020	14/08/2020
							F1905	Device Revision or Replacement	A072201	High Impedance	14/08/2020	14/08/2020
9837	2019	All other reportable incidents	47265	Dual-chamber implantable pacemaker, rate-responsive	CT141	Electrodes/Leads and associated devices	F1905	Device Revision or Replacement	A0701	Capturing Problem		
					CT145	Cardiovascular devices	F1905	Device Revision or Replacement	A0701	Capturing Problem		
					CT1684	Cardiac pacemakers and associated devices	F1905	Device Revision or Replacement	A0701	Capturing Problem		
					CT1685	Cardiac pacemakers	F1905	Device Revision or Replacement	A0701	Capturing Problem		
					CT233	Surgical	F1905	Device Revision or Replacement	A0701	Capturing Problem		
					CT2406	Implantable	F1905	Device Revision or Replacement	A0701	Capturing Problem		
					CT2407	Active implantable	F1905	Device Revision or Replacement	A0701	Capturing Problem		
					CT257	Implantable cardiac pacemakers	F1905	Device Revision or Replacement	A0701	Capturing Problem		
					CT2727	Active implant devices/ancillaries	F1905	Device Revision or Replacement	A0701	Capturing Problem		
					CT2750	Implantable cardiac electrical stimulators and associated devices	F1905	Device Revision or Replacement	A0701	Capturing Problem		
					CT2866	Implantable cardiac electrical stimulators	F1905	Device Revision or Replacement	A0701	Capturing Problem		
					CT2985	Active	F1905	Device Revision or Replacement	A0701	Capturing Problem		
					CT3001	Implantable metal	F1905	Device Revision or Replacement	A0701	Capturing Problem		
					CT3002	Implantable synthetic polymer	F1905	Device Revision or Replacement	A0701	Capturing Problem		
					CT3006	Implantable inorganic material	F1905	Device Revision or Replacement	A0701	Capturing Problem		
					CT314	Electrically powered	F1905	Device Revision or Replacement	A0701	Capturing Problem		
					CT321	Long-term surgical invasive	F1905	Device Revision or Replacement	A0701	Capturing Problem		
					CT449	Generators and associated devices	F1905	Device Revision or Replacement	A0701	Capturing Problem		
					CT981	Single use	F1905	Device Revision or Replacement	A0701	Capturing Problem		
					CT983	Surgical invasive	F1905	Device Revision or Replacement	A0701	Capturing Problem		
9838	2020	All other reportable incidents	35223	Endocardial/interventricular septal pacing lead	CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070909	Oversensing		
									A090801	Signal Artifact/Noise		
							F26	No Health Consequences or Impact	A070909	Oversensing		
									A090801	Signal Artifact/Noise		
9839	2020	All other reportable incidents	17882	Physiologic-monitoring defibrillation system	CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070803	Failure to Power Up		
							F27	Problem Identified During Non-Clinical Procedure	A070803	Failure to Power Up		
9840	2020	All other reportable incidents	17882	Physiologic-monitoring defibrillation system	CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070803	Failure to Power Up		
							F27	Problem Identified During Non-Clinical Procedure	A070803	Failure to Power Up		
9841	2020	All other reportable incidents	17882	Physiologic-monitoring defibrillation system	CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070803	Failure to Power Up		
							F27	Problem Identified During Non-Clinical Procedure	A070803	Failure to Power Up		
9842	2020	All other reportable incidents	17882	Physiologic-monitoring defibrillation system	CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070803	Failure to Power Up		
							F27	Problem Identified During Non-Clinical Procedure	A070803	Failure to Power Up		
9843	2020	All other reportable incidents	17882	Physiologic-monitoring defibrillation system	CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070803	Failure to Power Up		
							F27	Problem Identified During Non-Clinical Procedure	A070803	Failure to Power Up		
9844	2020	All other reportable incidents	17882	Physiologic-monitoring defibrillation system	CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070803	Failure to Power Up		
							F27	Problem Identified During Non-Clinical Procedure	A070803	Failure to Power Up		
9845	2020	All other reportable incidents	17882	Physiologic-monitoring defibrillation system	CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070803	Failure to Power Up		
							F27	Problem Identified During Non-Clinical Procedure	A070803	Failure to Power Up		
9846	2020	All other reportable incidents	17882	Physiologic-monitoring defibrillation system	CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070803	Failure to Power Up		
							F27	Problem Identified During Non-Clinical Procedure	A070803	Failure to Power Up		
9847	2020	All other reportable incidents	17882	Physiologic-monitoring defibrillation system	CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070803	Failure to Power Up		
							F27	Problem Identified During Non-Clinical Procedure	A070803	Failure to Power Up		
9848	2020	All other reportable incidents	17882	Physiologic-monitoring defibrillation system	CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070803	Failure to Power Up		
							F27	Problem Identified During Non-Clinical Procedure	A070803	Failure to Power Up		
9849	2020	All other reportable incidents	17882	Physiologic-monitoring defibrillation system	CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070803	Failure to Power Up		
							F27	Problem Identified During Non-Clinical Procedure	A070803	Failure to Power Up		
9850	2019	All other reportable incidents	36241	Pacemaker/defibrillator lead	CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A0701	Capturing Problem		
9851	2020	All other reportable incidents	17882	Physiologic-monitoring defibrillation system	CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070803	Failure to Power Up		
							F27	Problem Identified During Non-Clinical Procedure	A070803	Failure to Power Up		
9852	2020	All other reportable incidents	17882	Physiologic-monitoring defibrillation system	CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070803	Failure to Power Up		
							F27	Problem Identified During Non-Clinical Procedure	A070803	Failure to Power Up		
9853	2020	All other reportable incidents	17882	Physiologic-monitoring defibrillation system	CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070803	Failure to Power Up		
							F27	Problem Identified During Non-Clinical Procedure	A070803	Failure to Power Up		
9854	2020	All other reportable incidents	47265	Dual-chamber implantable pacemaker, rate-responsive	CT141	Electrodes/Leads and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	21/08/2020	21/08/2020
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	21/08/2020	21/08/2020
							F12	Serious Injury/ Illness/ Impairment	A070504	Premature Discharge of Battery	21/08/2020	21/08/2020
							F1903	Device Explantation	A070504	Premature Discharge of Battery	21/08/2020	21/08/2020
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	21/08/2020	21/08/2020
					CT145	Cardiovascular devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	21/08/2020	21/08/2020
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	21/08/2020	21/08/2020
							F12	Serious Injury/ Illness/ Impairment	A070504	Premature Discharge of Battery	21/08/2020	21/08/2020
							F1903	Device Explantation	A070504	Premature Discharge of Battery	21/08/2020	21/08/2020
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	21/08/2020	21/08/2020
					CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	21/08/2020	21/08/2020
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	21/08/2020	21/08/2020
							F12	Serious Injury/ Illness/ Impairment	A070504	Premature Discharge of Battery	21/08/2020	21/08/2020
							F1903	Device Explantation	A070504	Premature Discharge of Battery	21/08/2020	21/08/2020
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	21/08/2020	21/08/2020
					CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	21/08/2020	21/08/2020
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	21/08/2020	21/08/2020
							F12	Serious Injury/ Illness/ Impairment	A070504	Premature Discharge of Battery	21/08/2020	21/08/2020

							F1903	Device Explantation	A070504	Premature Discharge of Battery	21/08/2020	21/08/2020
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	21/08/2020	21/08/2020
					CT233	Surgical	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	21/08/2020	21/08/2020
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	21/08/2020	21/08/2020
							F12	Serious Injury/ Illness/ Impairment	A070504	Premature Discharge of Battery	21/08/2020	21/08/2020
							F1903	Device Explantation	A070504	Premature Discharge of Battery	21/08/2020	21/08/2020
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	21/08/2020	21/08/2020
					CT2406	Implantable	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	21/08/2020	21/08/2020
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	21/08/2020	21/08/2020
							F12	Serious Injury/ Illness/ Impairment	A070504	Premature Discharge of Battery	21/08/2020	21/08/2020
							F1903	Device Explantation	A070504	Premature Discharge of Battery	21/08/2020	21/08/2020
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	21/08/2020	21/08/2020
					CT2407	Active implantable	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	21/08/2020	21/08/2020
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	21/08/2020	21/08/2020
							F12	Serious Injury/ Illness/ Impairment	A070504	Premature Discharge of Battery	21/08/2020	21/08/2020
							F1903	Device Explantation	A070504	Premature Discharge of Battery	21/08/2020	21/08/2020
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	21/08/2020	21/08/2020
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	21/08/2020	21/08/2020
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	21/08/2020	21/08/2020
							F12	Serious Injury/ Illness/ Impairment	A070504	Premature Discharge of Battery	21/08/2020	21/08/2020
							F1903	Device Explantation	A070504	Premature Discharge of Battery	21/08/2020	21/08/2020
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	21/08/2020	21/08/2020
					CT2727	Active implant devices/ancillaries	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	21/08/2020	21/08/2020
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	21/08/2020	21/08/2020
							F12	Serious Injury/ Illness/ Impairment	A070504	Premature Discharge of Battery	21/08/2020	21/08/2020
							F1903	Device Explantation	A070504	Premature Discharge of Battery	21/08/2020	21/08/2020
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	21/08/2020	21/08/2020
					CT2750	Implantable cardiac electrical stimulators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	21/08/2020	21/08/2020
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	21/08/2020	21/08/2020
							F12	Serious Injury/ Illness/ Impairment	A070504	Premature Discharge of Battery	21/08/2020	21/08/2020
							F1903	Device Explantation	A070504	Premature Discharge of Battery	21/08/2020	21/08/2020
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	21/08/2020	21/08/2020
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	21/08/2020	21/08/2020
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	21/08/2020	21/08/2020
							F12	Serious Injury/ Illness/ Impairment	A070504	Premature Discharge of Battery	21/08/2020	21/08/2020
							F1903	Device Explantation	A070504	Premature Discharge of Battery	21/08/2020	21/08/2020
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	21/08/2020	21/08/2020
					CT2985	Active	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	21/08/2020	21/08/2020
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	21/08/2020	21/08/2020
							F12	Serious Injury/ Illness/ Impairment	A070504	Premature Discharge of Battery	21/08/2020	21/08/2020
							F1903	Device Explantation	A070504	Premature Discharge of Battery	21/08/2020	21/08/2020
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	21/08/2020	21/08/2020
					CT3001	Implantable metal	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	21/08/2020	21/08/2020
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	21/08/2020	21/08/2020
							F12	Serious Injury/ Illness/ Impairment	A070504	Premature Discharge of Battery	21/08/2020	21/08/2020
							F1903	Device Explantation	A070504	Premature Discharge of Battery	21/08/2020	21/08/2020
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	21/08/2020	21/08/2020
					CT3002	Implantable synthetic polymer	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	21/08/2020	21/08/2020
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	21/08/2020	21/08/2020
							F12	Serious Injury/ Illness/ Impairment	A070504	Premature Discharge of Battery	21/08/2020	21/08/2020
							F1903	Device Explantation	A070504	Premature Discharge of Battery	21/08/2020	21/08/2020
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	21/08/2020	21/08/2020
					CT3006	Implantable inorganic material	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	21/08/2020	21/08/2020
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	21/08/2020	21/08/2020
							F12	Serious Injury/ Illness/ Impairment	A070504	Premature Discharge of Battery	21/08/2020	21/08/2020
							F1903	Device Explantation	A070504	Premature Discharge of Battery	21/08/2020	21/08/2020
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	21/08/2020	21/08/2020
					CT314	Electrically powered	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	21/08/2020	21/08/2020
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	21/08/2020	21/08/2020
							F12	Serious Injury/ Illness/ Impairment	A070504	Premature Discharge of Battery	21/08/2020	21/08/2020
							F1903	Device Explantation	A070504	Premature Discharge of Battery	21/08/2020	21/08/2020
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	21/08/2020	21/08/2020
					CT321	Long-term surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	21/08/2020	21/08/2020
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	21/08/2020	21/08/2020
							F12	Serious Injury/ Illness/ Impairment	A070504	Premature Discharge of Battery	21/08/2020	21/08/2020
							F1903	Device Explantation	A070504	Premature Discharge of Battery	21/08/2020	21/08/2020
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	21/08/2020	21/08/2020
					CT449	Generators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	21/08/2020	21/08/2020
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	21/08/2020	21/08/2020
							F12	Serious Injury/ Illness/ Impairment	A070504	Premature Discharge of Battery	21/08/2020	21/08/2020
							F1903	Device Explantation	A070504	Premature Discharge of Battery	21/08/2020	21/08/2020
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	21/08/2020	21/08/2020
					CT981	Single use	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	21/08/2020	21/08/2020
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	21/08/2020	21/08/2020
							F12	Serious Injury/ Illness/ Impairment	A070504	Premature Discharge of Battery	21/08/2020	21/08/2020
							F1903	Device Explantation	A070504	Premature Discharge of Battery	21/08/2020	21/08/2020
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	21/08/2020	21/08/2020
					CT983	Surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	21/08/2020	21/08/2020
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	21/08/2020	21/08/2020
							F12	Serious Injury/ Illness/ Impairment	A070504	Premature Discharge of Battery	21/08/2020	21/08/2020
							F1903	Device Explantation	A070504	Premature Discharge of Battery	21/08/2020	21/08/2020
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	21/08/2020	21/08/2020
9855	2020	All other reportable incidents	47270	Cardiac resynchronization therapy implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	13/08/2020	13/08/2020
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	13/08/2020	13/08/2020
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery	13/08/2020	13/08/2020
9856	2020	All other reportable incidents	47265	Dual-chamber implantable pacemaker, rate-responsive	CT141	Electrodes/Leads and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A090804	False Positive Result		
							F26	No Health Consequences or Impact	A090804	False Positive Result		
					CT145	Cardiovascular devices	E2403	No Clinical Signs, Symptoms or Conditions	A090804	False Positive Result		

							F26	No Health Consequences or Impact	A090804	False Positive Result		
					CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A090804	False Positive Result		
							F26	No Health Consequences or Impact	A090804	False Positive Result		
					CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A090804	False Positive Result		
							F26	No Health Consequences or Impact	A090804	False Positive Result		
					CT233	Surgical	E2403	No Clinical Signs, Symptoms or Conditions	A090804	False Positive Result		
							F26	No Health Consequences or Impact	A090804	False Positive Result		
					CT2406	Implantable	E2403	No Clinical Signs, Symptoms or Conditions	A090804	False Positive Result		
							F26	No Health Consequences or Impact	A090804	False Positive Result		
					CT2407	Active implantable	E2403	No Clinical Signs, Symptoms or Conditions	A090804	False Positive Result		
							F26	No Health Consequences or Impact	A090804	False Positive Result		
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A090804	False Positive Result		
							F26	No Health Consequences or Impact	A090804	False Positive Result		
					CT2727	Active implant devices/ancillaries	E2403	No Clinical Signs, Symptoms or Conditions	A090804	False Positive Result		
							F26	No Health Consequences or Impact	A090804	False Positive Result		
					CT2750	Implantable cardiac electrical stimulators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A090804	False Positive Result		
							F26	No Health Consequences or Impact	A090804	False Positive Result		
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A090804	False Positive Result		
							F26	No Health Consequences or Impact	A090804	False Positive Result		
					CT2985	Active	E2403	No Clinical Signs, Symptoms or Conditions	A090804	False Positive Result		
							F26	No Health Consequences or Impact	A090804	False Positive Result		
					CT3001	Implantable metal	E2403	No Clinical Signs, Symptoms or Conditions	A090804	False Positive Result		
							F26	No Health Consequences or Impact	A090804	False Positive Result		
					CT3002	Implantable synthetic polymer	E2403	No Clinical Signs, Symptoms or Conditions	A090804	False Positive Result		
							F26	No Health Consequences or Impact	A090804	False Positive Result		
					CT3006	Implantable inorganic material	E2403	No Clinical Signs, Symptoms or Conditions	A090804	False Positive Result		
							F26	No Health Consequences or Impact	A090804	False Positive Result		
					CT314	Electrically powered	E2403	No Clinical Signs, Symptoms or Conditions	A090804	False Positive Result		
							F26	No Health Consequences or Impact	A090804	False Positive Result		
					CT321	Long-term surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A090804	False Positive Result		
							F26	No Health Consequences or Impact	A090804	False Positive Result		
					CT449	Generators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A090804	False Positive Result		
							F26	No Health Consequences or Impact	A090804	False Positive Result		
					CT981	Single use	E2403	No Clinical Signs, Symptoms or Conditions	A090804	False Positive Result		
							F26	No Health Consequences or Impact	A090804	False Positive Result		
					CT983	Surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A090804	False Positive Result		
							F26	No Health Consequences or Impact	A090804	False Positive Result		
9857	2019	All other reportable incidents	35852	Single-chamber implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	F1905	Device Revision or Replacement	A160105	False Alarm		
9858	2020	All other reportable incidents	35852	Single-chamber implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A071302	Inappropriate/Inadequate Shock/Stimulation		
							F26	No Health Consequences or Impact	A071302	Inappropriate/Inadequate Shock/Stimulation		
9859	2020	All other reportable incidents	35852	Single-chamber implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E2104	Electric Shock	A0712	Pacing Problem		
									A071302	Inappropriate/Inadequate Shock/Stimulation		
							E2403	No Clinical Signs, Symptoms or Conditions	A0712	Pacing Problem		
									A071302	Inappropriate/Inadequate Shock/Stimulation		
							F24	Insufficient Health Impact Information	A0712	Pacing Problem		
									A071302	Inappropriate/Inadequate Shock/Stimulation		
9860	2020	All other reportable incidents	36241	Pacemaker/defibrillator lead	CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A0203	Defective Device	10/11/2020	10/11/2020
									A070909	Oversensing	10/11/2020	10/11/2020
							F08	Hospitalization or Prolonged Hospitalization	A0203	Defective Device	10/11/2020	10/11/2020
									A070909	Oversensing	10/11/2020	10/11/2020
							F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A0203	Defective Device	10/11/2020	10/11/2020
									A070909	Oversensing	10/11/2020	10/11/2020
							F12	Serious Injury/ Illness/ Impairment	A0203	Defective Device	10/11/2020	10/11/2020
									A070909	Oversensing	10/11/2020	10/11/2020
							F19	Surgical Intervention	A0203	Defective Device	10/11/2020	10/11/2020
									A070909	Oversensing	10/11/2020	10/11/2020
9861	2020	All other reportable incidents	36241	Pacemaker/defibrillator lead	CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A072201	High Impedance	16/06/2020	16/06/2020
							F19	Surgical Intervention	A072201	High Impedance	16/06/2020	16/06/2020
9862	2020	All other reportable incidents	35852	Single-chamber implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A0908	Incorrect, Inadequate or Imprecise Result or Readings		
							F26	No Health Consequences or Impact	A0908	Incorrect, Inadequate or Imprecise Result or Readings		
9863	2020	All other reportable incidents	35852	Single-chamber implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
9864	2020	All other reportable incidents	37265	Dual-chamber implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A070909	Oversensing		
									A090801	Signal Artifact/Noise		
							F26	No Health Consequences or Impact	A070909	Oversensing		
									A090801	Signal Artifact/Noise		
9865	2020	All other reportable incidents	36241	Pacemaker/defibrillator lead	CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070909	Oversensing		
									A090801	Signal Artifact/Noise		
							F26	No Health Consequences or Impact	A070909	Oversensing		
									A090801	Signal Artifact/Noise		
9866	2020	All other reportable incidents	37265	Dual-chamber implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A070101	Failure to Capture		
									A070909	Oversensing		
									A1301	Failure to Read Input Signal		
							F15	Recognized Device or Procedural Complication	A070101	Failure to Capture		
									A070909	Oversensing		
									A1301	Failure to Read Input Signal		
							F26	No Health Consequences or Impact	A070101	Failure to Capture		
									A070909	Oversensing		
									A1301	Failure to Read Input Signal		
9867	2019	All other reportable incidents	47265	Dual-chamber implantable pacemaker, rate-responsive	CT141	Electrodes/Leads and associated devices	F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	20/09/2019	20/09/2019
					CT145	Cardiovascular devices	F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	20/09/2019	20/09/2019
					CT1684	Cardiac pacemakers and associated devices	F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	20/09/2019	20/09/2019
					CT1685	Cardiac pacemakers	F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	20/09/2019	20/09/2019
					CT233	Surgical	F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	20/09/2019	20/09/2019
					CT2406	Implantable	F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	20/09/2019	20/09/2019
					CT2407	Active implantable	F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	20/09/2019	20/09/2019
					CT257	Implantable cardiac pacemakers	F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	20/09/2019	20/09/2019

					CT2727	Active implant devices/ancillaries	F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	20/09/2019	20/09/2019
					CT2750	Implantable cardiac electrical stimulators and associated devices	F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	20/09/2019	20/09/2019
					CT2866	Implantable cardiac electrical stimulators	F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	20/09/2019	20/09/2019
					CT2985	Active	F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	20/09/2019	20/09/2019
					CT3001	Implantable metal	F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	20/09/2019	20/09/2019
					CT3002	Implantable synthetic polymer	F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	20/09/2019	20/09/2019
					CT3006	Implantable inorganic material	F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	20/09/2019	20/09/2019
					CT314	Electrically powered	F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	20/09/2019	20/09/2019
					CT321	Long-term surgical invasive	F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	20/09/2019	20/09/2019
					CT449	Generators and associated devices	F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	20/09/2019	20/09/2019
					CT981	Single use	F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	20/09/2019	20/09/2019
					CT983	Surgical invasive	F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	20/09/2019	20/09/2019
9868	2019	All other reportable incidents	47267	Single-chamber implantable pacemaker, rate-responsive	CT141	Electrodes/Leads and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
					CT145	Cardiovascular devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
					CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
					CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
					CT233	Surgical	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
					CT2406	Implantable	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
					CT2407	Active implantable	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
					CT2727	Active implant devices/ancillaries	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
					CT2750	Implantable cardiac electrical stimulators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
					CT2985	Active	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
					CT3001	Implantable metal	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
					CT3002	Implantable synthetic polymer	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
					CT3006	Implantable inorganic material	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
					CT314	Electrically powered	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
					CT321	Long-term surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
					CT449	Generators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
					CT981	Single use	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
					CT983	Surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
9869	2019	All other reportable incidents	35852	Single-chamber implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E2117	Twiddlers Syndrome	A071302	Inappropriate/Inadequate Shock/Stimulation		
							F08	Hospitalization or Prolonged Hospitalization	A071302	Inappropriate/Inadequate Shock/Stimulation		
							F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A071302	Inappropriate/Inadequate Shock/Stimulation		
							F19	Surgical Intervention	A071302	Inappropriate/Inadequate Shock/Stimulation		
							F1905	Device Revision or Replacement	A071302	Inappropriate/Inadequate Shock/Stimulation		
							F26	No Health Consequences or Impact	A071302	Inappropriate/Inadequate Shock/Stimulation		
9870	2019	All other reportable incidents	36241	Pacemaker/defibrillator lead	CT1684	Cardiac pacemakers and associated devices	F1905	Device Revision or Replacement	A071302	Inappropriate/Inadequate Shock/Stimulation	18/09/2019	
9871	2020	All other reportable incidents	47265	Dual-chamber implantable pacemaker, rate-responsive	CT141	Electrodes/Leads and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	27/08/2020	27/08/2020
									A071102	Failure to Interrogate	27/08/2020	27/08/2020
							F1901	Additional Surgery	A070504	Premature Discharge of Battery	27/08/2020	27/08/2020
									A071102	Failure to Interrogate	27/08/2020	27/08/2020
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	27/08/2020	27/08/2020
									A071102	Failure to Interrogate	27/08/2020	27/08/2020
					CT145	Cardiovascular devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	27/08/2020	27/08/2020
									A071102	Failure to Interrogate	27/08/2020	27/08/2020
							F1901	Additional Surgery	A070504	Premature Discharge of Battery	27/08/2020	27/08/2020
									A071102	Failure to Interrogate	27/08/2020	27/08/2020
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	27/08/2020	27/08/2020
									A071102	Failure to Interrogate	27/08/2020	27/08/2020
					CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	27/08/2020	27/08/2020
									A071102	Failure to Interrogate	27/08/2020	27/08/2020
							F1901	Additional Surgery	A070504	Premature Discharge of Battery	27/08/2020	27/08/2020
									A071102	Failure to Interrogate	27/08/2020	27/08/2020
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	27/08/2020	27/08/2020
									A071102	Failure to Interrogate	27/08/2020	27/08/2020
					CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	27/08/2020	27/08/2020
									A071102	Failure to Interrogate	27/08/2020	27/08/2020
							F1901	Additional Surgery	A070504	Premature Discharge of Battery	27/08/2020	27/08/2020
									A071102	Failure to Interrogate	27/08/2020	27/08/2020
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	27/08/2020	27/08/2020
									A071102	Failure to Interrogate	27/08/2020	27/08/2020
					CT233	Surgical	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	27/08/2020	27/08/2020
									A071102	Failure to Interrogate	27/08/2020	27/08/2020
							F1901	Additional Surgery	A070504	Premature Discharge of Battery	27/08/2020	27/08/2020
									A071102	Failure to Interrogate	27/08/2020	27/08/2020
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	27/08/2020	27/08/2020
									A071102	Failure to Interrogate	27/08/2020	27/08/2020
					CT2406	Implantable	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	27/08/2020	27/08/2020
									A071102	Failure to Interrogate	27/08/2020	27/08/2020
							F1901	Additional Surgery	A070504	Premature Discharge of Battery	27/08/2020	27/08/2020
									A071102	Failure to Interrogate	27/08/2020	27/08/2020
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	27/08/2020	27/08/2020
									A071102	Failure to Interrogate	27/08/2020	27/08/2020
					CT2407	Active implantable	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	27/08/2020	27/08/2020
									A071102	Failure to Interrogate	27/08/2020	27/08/2020
							F1901	Additional Surgery	A070504	Premature Discharge of Battery	27/08/2020	27/08/2020
									A071102	Failure to Interrogate	27/08/2020	27/08/2020
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	27/08/2020	27/08/2020
									A071102	Failure to Interrogate	27/08/2020	27/08/2020
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	27/08/2020	27/08/2020
									A071102	Failure to Interrogate	27/08/2020	27/08/2020
							F1901	Additional Surgery	A070504	Premature Discharge of Battery	27/08/2020	27/08/2020
									A071102	Failure to Interrogate	27/08/2020	27/08/2020
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	27/08/2020	27/08/2020

									A071102	Failure to Interrogate	27/08/2020	27/08/2020
					CT2727	Active implant devices/ancillaries	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	27/08/2020	27/08/2020
									A071102	Failure to Interrogate	27/08/2020	27/08/2020
							F1901	Additional Surgery	A070504	Premature Discharge of Battery	27/08/2020	27/08/2020
									A071102	Failure to Interrogate	27/08/2020	27/08/2020
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	27/08/2020	27/08/2020
									A071102	Failure to Interrogate	27/08/2020	27/08/2020
					CT2750	Implantable cardiac electrical stimulators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	27/08/2020	27/08/2020
									A071102	Failure to Interrogate	27/08/2020	27/08/2020
							F1901	Additional Surgery	A070504	Premature Discharge of Battery	27/08/2020	27/08/2020
									A071102	Failure to Interrogate	27/08/2020	27/08/2020
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	27/08/2020	27/08/2020
									A071102	Failure to Interrogate	27/08/2020	27/08/2020
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	27/08/2020	27/08/2020
									A071102	Failure to Interrogate	27/08/2020	27/08/2020
							F1901	Additional Surgery	A070504	Premature Discharge of Battery	27/08/2020	27/08/2020
									A071102	Failure to Interrogate	27/08/2020	27/08/2020
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	27/08/2020	27/08/2020
									A071102	Failure to Interrogate	27/08/2020	27/08/2020
					CT2985	Active	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	27/08/2020	27/08/2020
									A071102	Failure to Interrogate	27/08/2020	27/08/2020
							F1901	Additional Surgery	A070504	Premature Discharge of Battery	27/08/2020	27/08/2020
									A071102	Failure to Interrogate	27/08/2020	27/08/2020
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	27/08/2020	27/08/2020
									A071102	Failure to Interrogate	27/08/2020	27/08/2020
					CT3001	Implantable metal	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	27/08/2020	27/08/2020
									A071102	Failure to Interrogate	27/08/2020	27/08/2020
							F1901	Additional Surgery	A070504	Premature Discharge of Battery	27/08/2020	27/08/2020
									A071102	Failure to Interrogate	27/08/2020	27/08/2020
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	27/08/2020	27/08/2020
									A071102	Failure to Interrogate	27/08/2020	27/08/2020
					CT3002	Implantable synthetic polymer	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	27/08/2020	27/08/2020
									A071102	Failure to Interrogate	27/08/2020	27/08/2020
							F1901	Additional Surgery	A070504	Premature Discharge of Battery	27/08/2020	27/08/2020
									A071102	Failure to Interrogate	27/08/2020	27/08/2020
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	27/08/2020	27/08/2020
									A071102	Failure to Interrogate	27/08/2020	27/08/2020
					CT3006	Implantable inorganic material	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	27/08/2020	27/08/2020
									A071102	Failure to Interrogate	27/08/2020	27/08/2020
							F1901	Additional Surgery	A070504	Premature Discharge of Battery	27/08/2020	27/08/2020
									A071102	Failure to Interrogate	27/08/2020	27/08/2020
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	27/08/2020	27/08/2020
									A071102	Failure to Interrogate	27/08/2020	27/08/2020
					CT314	Electrically powered	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	27/08/2020	27/08/2020
									A071102	Failure to Interrogate	27/08/2020	27/08/2020
							F1901	Additional Surgery	A070504	Premature Discharge of Battery	27/08/2020	27/08/2020
									A071102	Failure to Interrogate	27/08/2020	27/08/2020
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	27/08/2020	27/08/2020
									A071102	Failure to Interrogate	27/08/2020	27/08/2020
					CT321	Long-term surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	27/08/2020	27/08/2020
									A071102	Failure to Interrogate	27/08/2020	27/08/2020
							F1901	Additional Surgery	A070504	Premature Discharge of Battery	27/08/2020	27/08/2020
									A071102	Failure to Interrogate	27/08/2020	27/08/2020
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	27/08/2020	27/08/2020
									A071102	Failure to Interrogate	27/08/2020	27/08/2020
					CT449	Generators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	27/08/2020	27/08/2020
									A071102	Failure to Interrogate	27/08/2020	27/08/2020
							F1901	Additional Surgery	A070504	Premature Discharge of Battery	27/08/2020	27/08/2020
									A071102	Failure to Interrogate	27/08/2020	27/08/2020
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	27/08/2020	27/08/2020
									A071102	Failure to Interrogate	27/08/2020	27/08/2020
					CT981	Single use	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	27/08/2020	27/08/2020
									A071102	Failure to Interrogate	27/08/2020	27/08/2020
							F1901	Additional Surgery	A070504	Premature Discharge of Battery	27/08/2020	27/08/2020
									A071102	Failure to Interrogate	27/08/2020	27/08/2020
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	27/08/2020	27/08/2020
									A071102	Failure to Interrogate	27/08/2020	27/08/2020
					CT983	Surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	27/08/2020	27/08/2020
									A071102	Failure to Interrogate	27/08/2020	27/08/2020
							F1901	Additional Surgery	A070504	Premature Discharge of Battery	27/08/2020	27/08/2020
									A071102	Failure to Interrogate	27/08/2020	27/08/2020
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	27/08/2020	27/08/2020
									A071102	Failure to Interrogate	27/08/2020	27/08/2020
9872	2020	All other reportable incidents	35852	Single-chamber implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	30/04/2020	30/04/2020
							F1901	Additional Surgery	A070504	Premature Discharge of Battery	30/04/2020	30/04/2020
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	30/04/2020	30/04/2020
9873	2020	All other reportable incidents	36241	Pacemaker/defibrillator lead	CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A040101	Fracture		
									A070101	Failure to Capture		
									A072201	High Impedance		
							F1905	Device Revision or Replacement	A040101	Fracture		
									A070101	Failure to Capture		
									A072201	High Impedance		
9874	2020	All other reportable incidents	36241	Pacemaker/defibrillator lead	CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A05	Mechanical Problem		
									A070102	High Capture Threshold		
									A22	Human-Device Interface Problem		
							F1905	Device Revision or Replacement	A05	Mechanical Problem		
									A070102	High Capture Threshold		
									A22	Human-Device Interface Problem		

							F1908	Prolonged Surgery	A05	Mechanical Problem		
									A070102	High Capture Threshold		
									A22	Human-Device Interface Problem		
							F26	No Health Consequences or Impact	A05	Mechanical Problem		
									A070102	High Capture Threshold		
									A22	Human-Device Interface Problem		
9875	2020	All other reportable incidents	36241	Pacemaker/defibrillator lead	CT1684	Cardiac pacemakers and associated devices	E2104	Electric Shock	A071302	Inappropriate/Inadequate Shock/Stimulation	21/08/2020	21/08/2020
							F1905	Device Revision or Replacement	A071302	Inappropriate/Inadequate Shock/Stimulation	21/08/2020	21/08/2020
9876	2020	All other reportable incidents	47265	Dual-chamber implantable pacemaker, rate-responsive	CT141	Electrodes/Leads and associated devices	E2401	Insufficient Information	A05	Mechanical Problem	21/08/2020	
									A0712	Pacing Problem	21/08/2020	
									A072201	High Impedance	21/08/2020	
							F1905	Device Revision or Replacement	A05	Mechanical Problem	21/08/2020	
									A0712	Pacing Problem	21/08/2020	
									A072201	High Impedance	21/08/2020	
					CT145	Cardiovascular devices	E2401	Insufficient Information	A05	Mechanical Problem	21/08/2020	
									A0712	Pacing Problem	21/08/2020	
									A072201	High Impedance	21/08/2020	
							F1905	Device Revision or Replacement	A05	Mechanical Problem	21/08/2020	
									A0712	Pacing Problem	21/08/2020	
									A072201	High Impedance	21/08/2020	
					CT1684	Cardiac pacemakers and associated devices	E2401	Insufficient Information	A05	Mechanical Problem	21/08/2020	
									A0712	Pacing Problem	21/08/2020	
									A072201	High Impedance	21/08/2020	
							F1905	Device Revision or Replacement	A05	Mechanical Problem	21/08/2020	
									A0712	Pacing Problem	21/08/2020	
									A072201	High Impedance	21/08/2020	
					CT1685	Cardiac pacemakers	E2401	Insufficient Information	A05	Mechanical Problem	21/08/2020	
									A0712	Pacing Problem	21/08/2020	
									A072201	High Impedance	21/08/2020	
							F1905	Device Revision or Replacement	A05	Mechanical Problem	21/08/2020	
									A0712	Pacing Problem	21/08/2020	
									A072201	High Impedance	21/08/2020	
					CT233	Surgical	E2401	Insufficient Information	A05	Mechanical Problem	21/08/2020	
									A0712	Pacing Problem	21/08/2020	
									A072201	High Impedance	21/08/2020	
							F1905	Device Revision or Replacement	A05	Mechanical Problem	21/08/2020	
									A0712	Pacing Problem	21/08/2020	
									A072201	High Impedance	21/08/2020	
					CT2406	Implantable	E2401	Insufficient Information	A05	Mechanical Problem	21/08/2020	
									A0712	Pacing Problem	21/08/2020	
									A072201	High Impedance	21/08/2020	
							F1905	Device Revision or Replacement	A05	Mechanical Problem	21/08/2020	
									A0712	Pacing Problem	21/08/2020	
									A072201	High Impedance	21/08/2020	
					CT2407	Active implantable	E2401	Insufficient Information	A05	Mechanical Problem	21/08/2020	
									A0712	Pacing Problem	21/08/2020	
									A072201	High Impedance	21/08/2020	
							F1905	Device Revision or Replacement	A05	Mechanical Problem	21/08/2020	
									A0712	Pacing Problem	21/08/2020	
									A072201	High Impedance	21/08/2020	
					CT257	Implantable cardiac pacemakers	E2401	Insufficient Information	A05	Mechanical Problem	21/08/2020	
									A0712	Pacing Problem	21/08/2020	
									A072201	High Impedance	21/08/2020	
							F1905	Device Revision or Replacement	A05	Mechanical Problem	21/08/2020	
									A0712	Pacing Problem	21/08/2020	
									A072201	High Impedance	21/08/2020	
					CT2727	Active implant devices/ancillaries	E2401	Insufficient Information	A05	Mechanical Problem	21/08/2020	
									A0712	Pacing Problem	21/08/2020	
									A072201	High Impedance	21/08/2020	
							F1905	Device Revision or Replacement	A05	Mechanical Problem	21/08/2020	
									A0712	Pacing Problem	21/08/2020	
									A072201	High Impedance	21/08/2020	
					CT2750	Implantable cardiac electrical stimulators and associated devices	E2401	Insufficient Information	A05	Mechanical Problem	21/08/2020	
									A0712	Pacing Problem	21/08/2020	
									A072201	High Impedance	21/08/2020	
							F1905	Device Revision or Replacement	A05	Mechanical Problem	21/08/2020	
									A0712	Pacing Problem	21/08/2020	
									A072201	High Impedance	21/08/2020	
					CT2866	Implantable cardiac electrical stimulators	E2401	Insufficient Information	A05	Mechanical Problem	21/08/2020	
									A0712	Pacing Problem	21/08/2020	
									A072201	High Impedance	21/08/2020	
							F1905	Device Revision or Replacement	A05	Mechanical Problem	21/08/2020	
									A0712	Pacing Problem	21/08/2020	
									A072201	High Impedance	21/08/2020	
					CT2985	Active	E2401	Insufficient Information	A05	Mechanical Problem	21/08/2020	
									A0712	Pacing Problem	21/08/2020	
									A072201	High Impedance	21/08/2020	
							F1905	Device Revision or Replacement	A05	Mechanical Problem	21/08/2020	
									A0712	Pacing Problem	21/08/2020	
									A072201	High Impedance	21/08/2020	
					CT3001	Implantable metal	E2401	Insufficient Information	A05	Mechanical Problem	21/08/2020	
									A0712	Pacing Problem	21/08/2020	
									A072201	High Impedance	21/08/2020	
							F1905	Device Revision or Replacement	A05	Mechanical Problem	21/08/2020	
									A0712	Pacing Problem	21/08/2020	
									A072201	High Impedance	21/08/2020	
					CT3002	Implantable synthetic polymer	E2401	Insufficient Information	A05	Mechanical Problem	21/08/2020	
									A0712	Pacing Problem	21/08/2020	

									A072201	High Impedance	21/08/2020	
							F1905	Device Revision or Replacement	A05	Mechanical Problem	21/08/2020	
									A0712	Pacing Problem	21/08/2020	
									A072201	High Impedance	21/08/2020	
					CT3006	Implantable inorganic material	E2401	Insufficient Information	A05	Mechanical Problem	21/08/2020	
									A0712	Pacing Problem	21/08/2020	
									A072201	High Impedance	21/08/2020	
							F1905	Device Revision or Replacement	A05	Mechanical Problem	21/08/2020	
									A0712	Pacing Problem	21/08/2020	
									A072201	High Impedance	21/08/2020	
					CT314	Electrically powered	E2401	Insufficient Information	A05	Mechanical Problem	21/08/2020	
									A0712	Pacing Problem	21/08/2020	
									A072201	High Impedance	21/08/2020	
							F1905	Device Revision or Replacement	A05	Mechanical Problem	21/08/2020	
									A0712	Pacing Problem	21/08/2020	
									A072201	High Impedance	21/08/2020	
					CT321	Long-term surgical invasive	E2401	Insufficient Information	A05	Mechanical Problem	21/08/2020	
									A0712	Pacing Problem	21/08/2020	
									A072201	High Impedance	21/08/2020	
							F1905	Device Revision or Replacement	A05	Mechanical Problem	21/08/2020	
									A0712	Pacing Problem	21/08/2020	
									A072201	High Impedance	21/08/2020	
					CT449	Generators and associated devices	E2401	Insufficient Information	A05	Mechanical Problem	21/08/2020	
									A0712	Pacing Problem	21/08/2020	
									A072201	High Impedance	21/08/2020	
							F1905	Device Revision or Replacement	A05	Mechanical Problem	21/08/2020	
									A0712	Pacing Problem	21/08/2020	
									A072201	High Impedance	21/08/2020	
					CT981	Single use	E2401	Insufficient Information	A05	Mechanical Problem	21/08/2020	
									A0712	Pacing Problem	21/08/2020	
									A072201	High Impedance	21/08/2020	
							F1905	Device Revision or Replacement	A05	Mechanical Problem	21/08/2020	
									A0712	Pacing Problem	21/08/2020	
									A072201	High Impedance	21/08/2020	
					CT983	Surgical invasive	E2401	Insufficient Information	A05	Mechanical Problem	21/08/2020	
									A0712	Pacing Problem	21/08/2020	
									A072201	High Impedance	21/08/2020	
							F1905	Device Revision or Replacement	A05	Mechanical Problem	21/08/2020	
									A0712	Pacing Problem	21/08/2020	
									A072201	High Impedance	21/08/2020	
9877	2020	All other reportable incidents	47265	Dual-chamber implantable pacemaker, rate-responsive	CT141	Electrodes/Leads and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	01/08/2020	01/08/2020
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	01/08/2020	01/08/2020
							F1903	Device Explantation	A070504	Premature Discharge of Battery	01/08/2020	01/08/2020
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	01/08/2020	01/08/2020
					CT145	Cardiovascular devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	01/08/2020	01/08/2020
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	01/08/2020	01/08/2020
							F1903	Device Explantation	A070504	Premature Discharge of Battery	01/08/2020	01/08/2020
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	01/08/2020	01/08/2020
					CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	01/08/2020	01/08/2020
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	01/08/2020	01/08/2020
							F1903	Device Explantation	A070504	Premature Discharge of Battery	01/08/2020	01/08/2020
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	01/08/2020	01/08/2020
					CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	01/08/2020	01/08/2020
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	01/08/2020	01/08/2020
							F1903	Device Explantation	A070504	Premature Discharge of Battery	01/08/2020	01/08/2020
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	01/08/2020	01/08/2020
					CT233	Surgical	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	01/08/2020	01/08/2020
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	01/08/2020	01/08/2020
							F1903	Device Explantation	A070504	Premature Discharge of Battery	01/08/2020	01/08/2020
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	01/08/2020	01/08/2020
					CT2406	Implantable	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	01/08/2020	01/08/2020
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	01/08/2020	01/08/2020
							F1903	Device Explantation	A070504	Premature Discharge of Battery	01/08/2020	01/08/2020
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	01/08/2020	01/08/2020
					CT2407	Active implantable	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	01/08/2020	01/08/2020
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	01/08/2020	01/08/2020
							F1903	Device Explantation	A070504	Premature Discharge of Battery	01/08/2020	01/08/2020
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	01/08/2020	01/08/2020
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	01/08/2020	01/08/2020
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	01/08/2020	01/08/2020
							F1903	Device Explantation	A070504	Premature Discharge of Battery	01/08/2020	01/08/2020
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	01/08/2020	01/08/2020
					CT2727	Active implant devices/ancillaries	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	01/08/2020	01/08/2020
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	01/08/2020	01/08/2020
							F1903	Device Explantation	A070504	Premature Discharge of Battery	01/08/2020	01/08/2020
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	01/08/2020	01/08/2020
					CT2750	Implantable cardiac electrical stimulators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	01/08/2020	01/08/2020
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	01/08/2020	01/08/2020
							F1903	Device Explantation	A070504	Premature Discharge of Battery	01/08/2020	01/08/2020
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	01/08/2020	01/08/2020
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	01/08/2020	01/08/2020
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	01/08/2020	01/08/2020
							F1903	Device Explantation	A070504	Premature Discharge of Battery	01/08/2020	01/08/2020
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	01/08/2020	01/08/2020
					CT2985	Active	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	01/08/2020	01/08/2020
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	01/08/2020	01/08/2020
							F1903	Device Explantation	A070504	Premature Discharge of Battery	01/08/2020	01/08/2020
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	01/08/2020	01/08/2020

					CT3001	Implantable metal	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	01/08/2020	01/08/2020
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	01/08/2020	01/08/2020
							F1903	Device Explantation	A070504	Premature Discharge of Battery	01/08/2020	01/08/2020
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	01/08/2020	01/08/2020
					CT3002	Implantable synthetic polymer	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	01/08/2020	01/08/2020
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	01/08/2020	01/08/2020
							F1903	Device Explantation	A070504	Premature Discharge of Battery	01/08/2020	01/08/2020
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	01/08/2020	01/08/2020
					CT3006	Implantable inorganic material	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	01/08/2020	01/08/2020
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	01/08/2020	01/08/2020
							F1903	Device Explantation	A070504	Premature Discharge of Battery	01/08/2020	01/08/2020
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	01/08/2020	01/08/2020
					CT314	Electrically powered	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	01/08/2020	01/08/2020
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	01/08/2020	01/08/2020
							F1903	Device Explantation	A070504	Premature Discharge of Battery	01/08/2020	01/08/2020
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	01/08/2020	01/08/2020
					CT321	Long-term surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	01/08/2020	01/08/2020
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	01/08/2020	01/08/2020
							F1903	Device Explantation	A070504	Premature Discharge of Battery	01/08/2020	01/08/2020
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	01/08/2020	01/08/2020
					CT449	Generators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	01/08/2020	01/08/2020
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	01/08/2020	01/08/2020
							F1903	Device Explantation	A070504	Premature Discharge of Battery	01/08/2020	01/08/2020
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	01/08/2020	01/08/2020
					CT981	Single use	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	01/08/2020	01/08/2020
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	01/08/2020	01/08/2020
							F1903	Device Explantation	A070504	Premature Discharge of Battery	01/08/2020	01/08/2020
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	01/08/2020	01/08/2020
					CT983	Surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	01/08/2020	01/08/2020
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	01/08/2020	01/08/2020
							F1903	Device Explantation	A070504	Premature Discharge of Battery	01/08/2020	01/08/2020
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	01/08/2020	01/08/2020
9878	2019	All other reportable incidents	47267	Single-chamber implantable pacemaker, rate-responsive	CT141	Electrodes/Leads and associated devices	E1906	Unspecified Infection	A070101	Failure to Capture		
							F1905	Device Revision or Replacement	A070101	Failure to Capture		
					CT145	Cardiovascular devices	E1906	Unspecified Infection	A070101	Failure to Capture		
							F1905	Device Revision or Replacement	A070101	Failure to Capture		
					CT1684	Cardiac pacemakers and associated devices	E1906	Unspecified Infection	A070101	Failure to Capture		
							F1905	Device Revision or Replacement	A070101	Failure to Capture		
					CT1685	Cardiac pacemakers	E1906	Unspecified Infection	A070101	Failure to Capture		
							F1905	Device Revision or Replacement	A070101	Failure to Capture		
					CT233	Surgical	E1906	Unspecified Infection	A070101	Failure to Capture		
							F1905	Device Revision or Replacement	A070101	Failure to Capture		
					CT2406	Implantable	E1906	Unspecified Infection	A070101	Failure to Capture		
							F1905	Device Revision or Replacement	A070101	Failure to Capture		
					CT2407	Active implantable	E1906	Unspecified Infection	A070101	Failure to Capture		
							F1905	Device Revision or Replacement	A070101	Failure to Capture		
					CT257	Implantable cardiac pacemakers	E1906	Unspecified Infection	A070101	Failure to Capture		
							F1905	Device Revision or Replacement	A070101	Failure to Capture		
					CT2727	Active implant devices/ancillaries	E1906	Unspecified Infection	A070101	Failure to Capture		
							F1905	Device Revision or Replacement	A070101	Failure to Capture		
					CT2750	Implantable cardiac electrical stimulators and associated devices	E1906	Unspecified Infection	A070101	Failure to Capture		
							F1905	Device Revision or Replacement	A070101	Failure to Capture		
					CT2866	Implantable cardiac electrical stimulators	E1906	Unspecified Infection	A070101	Failure to Capture		
							F1905	Device Revision or Replacement	A070101	Failure to Capture		
					CT2985	Active	E1906	Unspecified Infection	A070101	Failure to Capture		
							F1905	Device Revision or Replacement	A070101	Failure to Capture		
					CT3001	Implantable metal	E1906	Unspecified Infection	A070101	Failure to Capture		
							F1905	Device Revision or Replacement	A070101	Failure to Capture		
					CT3002	Implantable synthetic polymer	E1906	Unspecified Infection	A070101	Failure to Capture		
							F1905	Device Revision or Replacement	A070101	Failure to Capture		
					CT3006	Implantable inorganic material	E1906	Unspecified Infection	A070101	Failure to Capture		
							F1905	Device Revision or Replacement	A070101	Failure to Capture		
					CT314	Electrically powered	E1906	Unspecified Infection	A070101	Failure to Capture		
							F1905	Device Revision or Replacement	A070101	Failure to Capture		
					CT321	Long-term surgical invasive	E1906	Unspecified Infection	A070101	Failure to Capture		
							F1905	Device Revision or Replacement	A070101	Failure to Capture		
					CT449	Generators and associated devices	E1906	Unspecified Infection	A070101	Failure to Capture		
							F1905	Device Revision or Replacement	A070101	Failure to Capture		
					CT981	Single use	E1906	Unspecified Infection	A070101	Failure to Capture		
							F1905	Device Revision or Replacement	A070101	Failure to Capture		
					CT983	Surgical invasive	E1906	Unspecified Infection	A070101	Failure to Capture		
							F1905	Device Revision or Replacement	A070101	Failure to Capture		
9879	2020	All other reportable incidents	37265	Dual-chamber implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E2401	Insufficient Information	A160601	Premature Elective Replacement Indicator		
							E2403	No Clinical Signs, Symptoms or Conditions	A160601	Premature Elective Replacement Indicator		
							F24	Insufficient Health Impact Information	A160601	Premature Elective Replacement Indicator		
9880	2020	All other reportable incidents	37265	Dual-chamber implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A160602	Premature End-of-Life Indicator		
							F1903	Device Explantation	A160602	Premature End-of-Life Indicator		
9881	2020	All other reportable incidents	47265	Dual-chamber implantable pacemaker, rate-responsive	CT141	Electrodes/Leads and associated devices	E060110	Ventricular Fibrillation	A0709	Device Sensing Problem		
									A0712	Pacing Problem		
									A090811	Unable to Obtain Readings		
							E0602	Cardiac Arrest	A0709	Device Sensing Problem		
									A0712	Pacing Problem		
									A090811	Unable to Obtain Readings		
							F1203	Life Threatening Illness or Injury	A0709	Device Sensing Problem		
									A0712	Pacing Problem		
									A090811	Unable to Obtain Readings		
					CT145	Cardiovascular devices	E060110	Ventricular Fibrillation	A0709	Device Sensing Problem		
									A0712	Pacing Problem		

								A090811	Unable to Obtain Readings		
						E0602	Cardiac Arrest	A0709	Device Sensing Problem		
								A0712	Pacing Problem		
								A090811	Unable to Obtain Readings		
						F1203	Life Threatening Illness or Injury	A0709	Device Sensing Problem		
								A0712	Pacing Problem		
								A090811	Unable to Obtain Readings		
					CT1684	Cardiac pacemakers and associated devices	E060110	Ventricular Fibrillation	A0709	Device Sensing Problem		
									A0712	Pacing Problem		
									A090811	Unable to Obtain Readings		
							E0602	Cardiac Arrest	A0709	Device Sensing Problem		
									A0712	Pacing Problem		
									A090811	Unable to Obtain Readings		
							F1203	Life Threatening Illness or Injury	A0709	Device Sensing Problem		
									A0712	Pacing Problem		
									A090811	Unable to Obtain Readings		
					CT1685	Cardiac pacemakers	E060110	Ventricular Fibrillation	A0709	Device Sensing Problem		
									A0712	Pacing Problem		
									A090811	Unable to Obtain Readings		
							E0602	Cardiac Arrest	A0709	Device Sensing Problem		
									A0712	Pacing Problem		
									A090811	Unable to Obtain Readings		
							F1203	Life Threatening Illness or Injury	A0709	Device Sensing Problem		
									A0712	Pacing Problem		
									A090811	Unable to Obtain Readings		
					CT233	Surgical	E060110	Ventricular Fibrillation	A0709	Device Sensing Problem		
									A0712	Pacing Problem		
									A090811	Unable to Obtain Readings		
							E0602	Cardiac Arrest	A0709	Device Sensing Problem		
									A0712	Pacing Problem		
									A090811	Unable to Obtain Readings		
							F1203	Life Threatening Illness or Injury	A0709	Device Sensing Problem		
									A0712	Pacing Problem		
									A090811	Unable to Obtain Readings		
					CT2406	Implantable	E060110	Ventricular Fibrillation	A0709	Device Sensing Problem		
									A0712	Pacing Problem		
									A090811	Unable to Obtain Readings		
							E0602	Cardiac Arrest	A0709	Device Sensing Problem		
									A0712	Pacing Problem		
									A090811	Unable to Obtain Readings		
							F1203	Life Threatening Illness or Injury	A0709	Device Sensing Problem		
									A0712	Pacing Problem		
									A090811	Unable to Obtain Readings		
					CT2407	Active implantable	E060110	Ventricular Fibrillation	A0709	Device Sensing Problem		
									A0712	Pacing Problem		
									A090811	Unable to Obtain Readings		
							E0602	Cardiac Arrest	A0709	Device Sensing Problem		
									A0712	Pacing Problem		
									A090811	Unable to Obtain Readings		
							F1203	Life Threatening Illness or Injury	A0709	Device Sensing Problem		
									A0712	Pacing Problem		
									A090811	Unable to Obtain Readings		
					CT257	Implantable cardiac pacemakers	E060110	Ventricular Fibrillation	A0709	Device Sensing Problem		
									A0712	Pacing Problem		
									A090811	Unable to Obtain Readings		
							E0602	Cardiac Arrest	A0709	Device Sensing Problem		
									A0712	Pacing Problem		
									A090811	Unable to Obtain Readings		
							F1203	Life Threatening Illness or Injury	A0709	Device Sensing Problem		
									A0712	Pacing Problem		
									A090811	Unable to Obtain Readings		
					CT2727	Active implant devices/ancillaries	E060110	Ventricular Fibrillation	A0709	Device Sensing Problem		
									A0712	Pacing Problem		
									A090811	Unable to Obtain Readings		
							E0602	Cardiac Arrest	A0709	Device Sensing Problem		
									A0712	Pacing Problem		
									A090811	Unable to Obtain Readings		
							F1203	Life Threatening Illness or Injury	A0709	Device Sensing Problem		
									A0712	Pacing Problem		
									A090811	Unable to Obtain Readings		
					CT2750	Implantable cardiac electrical stimulators and associated devices	E060110	Ventricular Fibrillation	A0709	Device Sensing Problem		
									A0712	Pacing Problem		
									A090811	Unable to Obtain Readings		
							E0602	Cardiac Arrest	A0709	Device Sensing Problem		
									A0712	Pacing Problem		
									A090811	Unable to Obtain Readings		
							F1203	Life Threatening Illness or Injury	A0709	Device Sensing Problem		
									A0712	Pacing Problem		
									A090811	Unable to Obtain Readings		
					CT2866	Implantable cardiac electrical stimulators	E060110	Ventricular Fibrillation	A0709	Device Sensing Problem		
									A0712	Pacing Problem		
									A090811	Unable to Obtain Readings		
							E0602	Cardiac Arrest	A0709	Device Sensing Problem		
									A0712	Pacing Problem		
									A090811	Unable to Obtain Readings		
							F1203	Life Threatening Illness or Injury	A0709	Device Sensing Problem		
									A0712	Pacing Problem		
									A090811	Unable to Obtain Readings		

					CT2985	Active	E060110	Ventricular Fibrillation	A0709	Device Sensing Problem		
									A0712	Pacing Problem		
									A090811	Unable to Obtain Readings		
							E0602	Cardiac Arrest	A0709	Device Sensing Problem		
									A0712	Pacing Problem		
									A090811	Unable to Obtain Readings		
							F1203	Life Threatening Illness or Injury	A0709	Device Sensing Problem		
									A0712	Pacing Problem		
									A090811	Unable to Obtain Readings		
					CT3001	Implantable metal	E060110	Ventricular Fibrillation	A0709	Device Sensing Problem		
									A0712	Pacing Problem		
									A090811	Unable to Obtain Readings		
							E0602	Cardiac Arrest	A0709	Device Sensing Problem		
									A0712	Pacing Problem		
									A090811	Unable to Obtain Readings		
							F1203	Life Threatening Illness or Injury	A0709	Device Sensing Problem		
									A0712	Pacing Problem		
									A090811	Unable to Obtain Readings		
					CT3002	Implantable synthetic polymer	E060110	Ventricular Fibrillation	A0709	Device Sensing Problem		
									A0712	Pacing Problem		
									A090811	Unable to Obtain Readings		
							E0602	Cardiac Arrest	A0709	Device Sensing Problem		
									A0712	Pacing Problem		
									A090811	Unable to Obtain Readings		
							F1203	Life Threatening Illness or Injury	A0709	Device Sensing Problem		
									A0712	Pacing Problem		
									A090811	Unable to Obtain Readings		
					CT3006	Implantable inorganic material	E060110	Ventricular Fibrillation	A0709	Device Sensing Problem		
									A0712	Pacing Problem		
									A090811	Unable to Obtain Readings		
							E0602	Cardiac Arrest	A0709	Device Sensing Problem		
									A0712	Pacing Problem		
									A090811	Unable to Obtain Readings		
							F1203	Life Threatening Illness or Injury	A0709	Device Sensing Problem		
									A0712	Pacing Problem		
									A090811	Unable to Obtain Readings		
					CT314	Electrically powered	E060110	Ventricular Fibrillation	A0709	Device Sensing Problem		
									A0712	Pacing Problem		
									A090811	Unable to Obtain Readings		
							E0602	Cardiac Arrest	A0709	Device Sensing Problem		
									A0712	Pacing Problem		
									A090811	Unable to Obtain Readings		
							F1203	Life Threatening Illness or Injury	A0709	Device Sensing Problem		
									A0712	Pacing Problem		
									A090811	Unable to Obtain Readings		
					CT321	Long-term surgical invasive	E060110	Ventricular Fibrillation	A0709	Device Sensing Problem		
									A0712	Pacing Problem		
									A090811	Unable to Obtain Readings		
							E0602	Cardiac Arrest	A0709	Device Sensing Problem		
									A0712	Pacing Problem		
									A090811	Unable to Obtain Readings		
							F1203	Life Threatening Illness or Injury	A0709	Device Sensing Problem		
									A0712	Pacing Problem		
									A090811	Unable to Obtain Readings		
					CT449	Generators and associated devices	E060110	Ventricular Fibrillation	A0709	Device Sensing Problem		
									A0712	Pacing Problem		
									A090811	Unable to Obtain Readings		
							E0602	Cardiac Arrest	A0709	Device Sensing Problem		
									A0712	Pacing Problem		
									A090811	Unable to Obtain Readings		
							F1203	Life Threatening Illness or Injury	A0709	Device Sensing Problem		
									A0712	Pacing Problem		
									A090811	Unable to Obtain Readings		
					CT981	Single use	E060110	Ventricular Fibrillation	A0709	Device Sensing Problem		
									A0712	Pacing Problem		
									A090811	Unable to Obtain Readings		
							E0602	Cardiac Arrest	A0709	Device Sensing Problem		
									A0712	Pacing Problem		
									A090811	Unable to Obtain Readings		
							F1203	Life Threatening Illness or Injury	A0709	Device Sensing Problem		
									A0712	Pacing Problem		
									A090811	Unable to Obtain Readings		
					CT983	Surgical invasive	E060110	Ventricular Fibrillation	A0709	Device Sensing Problem		
									A0712	Pacing Problem		
									A090811	Unable to Obtain Readings		
							E0602	Cardiac Arrest	A0709	Device Sensing Problem		
									A0712	Pacing Problem		
									A090811	Unable to Obtain Readings		
							F1203	Life Threatening Illness or Injury	A0709	Device Sensing Problem		
									A0712	Pacing Problem		
									A090811	Unable to Obtain Readings		
9882	2020	All other reportable incidents	47267	Single-chamber implantable pacemaker, rate-responsive	CT141	Electrodes/Leads and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A0709	Device Sensing Problem		
									A090404	Output Above Specifications		
									A090811	Unable to Obtain Readings		
							F26	No Health Consequences or Impact	A0709	Device Sensing Problem		
									A090404	Output Above Specifications		
									A090811	Unable to Obtain Readings		
					CT145	Cardiovascular devices	E2403	No Clinical Signs, Symptoms or Conditions	A0709	Device Sensing Problem		

									A090404	Output Above Specifications			
									A090811	Unable to Obtain Readings			
								F26	No Health Consequences or Impact	A0709	Device Sensing Problem		
										A090404	Output Above Specifications		
										A090811	Unable to Obtain Readings		
					CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A0709	Device Sensing Problem			
									A090404	Output Above Specifications			
									A090811	Unable to Obtain Readings			
								F26	No Health Consequences or Impact	A0709	Device Sensing Problem		
										A090404	Output Above Specifications		
										A090811	Unable to Obtain Readings		
					CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A0709	Device Sensing Problem			
									A090404	Output Above Specifications			
									A090811	Unable to Obtain Readings			
								F26	No Health Consequences or Impact	A0709	Device Sensing Problem		
										A090404	Output Above Specifications		
										A090811	Unable to Obtain Readings		
					CT233	Surgical	E2403	No Clinical Signs, Symptoms or Conditions	A0709	Device Sensing Problem			
									A090404	Output Above Specifications			
									A090811	Unable to Obtain Readings			
								F26	No Health Consequences or Impact	A0709	Device Sensing Problem		
										A090404	Output Above Specifications		
										A090811	Unable to Obtain Readings		
					CT2406	Implantable	E2403	No Clinical Signs, Symptoms or Conditions	A0709	Device Sensing Problem			
									A090404	Output Above Specifications			
									A090811	Unable to Obtain Readings			
								F26	No Health Consequences or Impact	A0709	Device Sensing Problem		
										A090404	Output Above Specifications		
										A090811	Unable to Obtain Readings		
					CT2407	Active implantable	E2403	No Clinical Signs, Symptoms or Conditions	A0709	Device Sensing Problem			
									A090404	Output Above Specifications			
									A090811	Unable to Obtain Readings			
								F26	No Health Consequences or Impact	A0709	Device Sensing Problem		
										A090404	Output Above Specifications		
										A090811	Unable to Obtain Readings		
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A0709	Device Sensing Problem			
									A090404	Output Above Specifications			
									A090811	Unable to Obtain Readings			
								F26	No Health Consequences or Impact	A0709	Device Sensing Problem		
										A090404	Output Above Specifications		
										A090811	Unable to Obtain Readings		
					CT2727	Active implant devices/ancillaries	E2403	No Clinical Signs, Symptoms or Conditions	A0709	Device Sensing Problem			
									A090404	Output Above Specifications			
									A090811	Unable to Obtain Readings			
								F26	No Health Consequences or Impact	A0709	Device Sensing Problem		
										A090404	Output Above Specifications		
										A090811	Unable to Obtain Readings		
					CT2750	Implantable cardiac electrical stimulators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A0709	Device Sensing Problem			
									A090404	Output Above Specifications			
									A090811	Unable to Obtain Readings			
								F26	No Health Consequences or Impact	A0709	Device Sensing Problem		
										A090404	Output Above Specifications		
										A090811	Unable to Obtain Readings		
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A0709	Device Sensing Problem			
									A090404	Output Above Specifications			
									A090811	Unable to Obtain Readings			
								F26	No Health Consequences or Impact	A0709	Device Sensing Problem		
										A090404	Output Above Specifications		
										A090811	Unable to Obtain Readings		
					CT2985	Active	E2403	No Clinical Signs, Symptoms or Conditions	A0709	Device Sensing Problem			
									A090404	Output Above Specifications			
									A090811	Unable to Obtain Readings			
								F26	No Health Consequences or Impact	A0709	Device Sensing Problem		
										A090404	Output Above Specifications		
										A090811	Unable to Obtain Readings		
					CT3001	Implantable metal	E2403	No Clinical Signs, Symptoms or Conditions	A0709	Device Sensing Problem			
									A090404	Output Above Specifications			
									A090811	Unable to Obtain Readings			
								F26	No Health Consequences or Impact	A0709	Device Sensing Problem		
										A090404	Output Above Specifications		
										A090811	Unable to Obtain Readings		
					CT3002	Implantable synthetic polymer	E2403	No Clinical Signs, Symptoms or Conditions	A0709	Device Sensing Problem			
									A090404	Output Above Specifications			
									A090811	Unable to Obtain Readings			
								F26	No Health Consequences or Impact	A0709	Device Sensing Problem		
										A090404	Output Above Specifications		
										A090811	Unable to Obtain Readings		
					CT3006	Implantable inorganic material	E2403	No Clinical Signs, Symptoms or Conditions	A0709	Device Sensing Problem			
									A090404	Output Above Specifications			
									A090811	Unable to Obtain Readings			
								F26	No Health Consequences or Impact	A0709	Device Sensing Problem		
										A090404	Output Above Specifications		
										A090811	Unable to Obtain Readings		
					CT314	Electrically powered	E2403	No Clinical Signs, Symptoms or Conditions	A0709	Device Sensing Problem			
									A090404	Output Above Specifications			
									A090811	Unable to Obtain Readings			
								F26	No Health Consequences or Impact	A0709	Device Sensing Problem		
										A090404	Output Above Specifications		

									A090811	Unable to Obtain Readings		
					CT321	Long-term surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A0709	Device Sensing Problem		
									A090404	Output Above Specifications		
									A090811	Unable to Obtain Readings		
							F26	No Health Consequences or Impact	A0709	Device Sensing Problem		
									A090404	Output Above Specifications		
									A090811	Unable to Obtain Readings		
					CT449	Generators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A0709	Device Sensing Problem		
									A090404	Output Above Specifications		
									A090811	Unable to Obtain Readings		
							F26	No Health Consequences or Impact	A0709	Device Sensing Problem		
									A090404	Output Above Specifications		
									A090811	Unable to Obtain Readings		
					CT981	Single use	E2403	No Clinical Signs, Symptoms or Conditions	A0709	Device Sensing Problem		
									A090404	Output Above Specifications		
									A090811	Unable to Obtain Readings		
							F26	No Health Consequences or Impact	A0709	Device Sensing Problem		
									A090404	Output Above Specifications		
									A090811	Unable to Obtain Readings		
					CT983	Surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A0709	Device Sensing Problem		
									A090404	Output Above Specifications		
									A090811	Unable to Obtain Readings		
							F26	No Health Consequences or Impact	A0709	Device Sensing Problem		
									A090404	Output Above Specifications		
									A090811	Unable to Obtain Readings		
9883	2020	All other reportable incidents	17882	Physiologic-monitoring defibrillation system	CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A0719	Unexpected Shutdown		
							F27	Problem Identified During Non-Clinical Procedure	A0719	Unexpected Shutdown		
9884	2020	All other reportable incidents	17882	Physiologic-monitoring defibrillation system	CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A0719	Unexpected Shutdown		
							F27	Problem Identified During Non-Clinical Procedure	A0719	Unexpected Shutdown		
9885	2020	All other reportable incidents	36241	Pacemaker/defibrillator lead	CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A090801	Signal Artifact/Noise		
									A1301	Failure to Read Input Signal		
							F19	Surgical Intervention	A090801	Signal Artifact/Noise		
									A1301	Failure to Read Input Signal		
							F1905	Device Revision or Replacement	A090801	Signal Artifact/Noise		
									A1301	Failure to Read Input Signal		
9886	2020	All other reportable incidents	35852	Single-chamber implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E2104	Electric Shock	A0712	Pacing Problem		
									A071302	Inappropriate/Inadequate Shock/Stimulation		
							E2403	No Clinical Signs, Symptoms or Conditions	A0712	Pacing Problem		
									A071302	Inappropriate/Inadequate Shock/Stimulation		
							F26	No Health Consequences or Impact	A0712	Pacing Problem		
									A071302	Inappropriate/Inadequate Shock/Stimulation		
9887	2020	All other reportable incidents	36241	Pacemaker/defibrillator lead	CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A051201	Device Dislodged or Dislocated		
									A070910	Undersensing		
									A0712	Pacing Problem		
							F08	Hospitalization or Prolonged Hospitalization	A051201	Device Dislodged or Dislocated		
									A070910	Undersensing		
									A0712	Pacing Problem		
							F1904	Device Repositioning	A051201	Device Dislodged or Dislocated		
									A070910	Undersensing		
									A0712	Pacing Problem		
9888	2020	All other reportable incidents	36241	Pacemaker/defibrillator lead	CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A0203	Defective Device	30/07/2020	30/07/2020
									A072201	High Impedance	30/07/2020	30/07/2020
							F05	Delay to Treatment/ Therapy	A0203	Defective Device	30/07/2020	30/07/2020
									A072201	High Impedance	30/07/2020	30/07/2020
							F1905	Device Revision or Replacement	A0203	Defective Device	30/07/2020	30/07/2020
									A072201	High Impedance	30/07/2020	30/07/2020
							F1908	Prolonged Surgery	A0203	Defective Device	30/07/2020	30/07/2020
									A072201	High Impedance	30/07/2020	30/07/2020
9889	2019	All other reportable incidents	47270	Cardiac resynchronization therapy implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	F1905	Device Revision or Replacement	A0705	Battery Problem	30/09/2019	
9890	2020	All other reportable incidents	47265	Dual-chamber implantable pacemaker, rate-responsive	CT141	Electrodes/Leads and associated devices	E2402	Appropriate Clinical Signs, Symptoms and Conditions Term/Code not Available	A0908	Incorrect, Inadequate or Imprecise Result or Readings		
							E2403	No Clinical Signs, Symptoms or Conditions	A0908	Incorrect, Inadequate or Imprecise Result or Readings		
							F24	Insufficient Health Impact Information	A0908	Incorrect, Inadequate or Imprecise Result or Readings		
					CT145	Cardiovascular devices	E2402	Appropriate Clinical Signs, Symptoms and Conditions Term/Code not Available	A0908	Incorrect, Inadequate or Imprecise Result or Readings		
							E2403	No Clinical Signs, Symptoms or Conditions	A0908	Incorrect, Inadequate or Imprecise Result or Readings		
							F24	Insufficient Health Impact Information	A0908	Incorrect, Inadequate or Imprecise Result or Readings		
					CT1684	Cardiac pacemakers and associated devices	E2402	Appropriate Clinical Signs, Symptoms and Conditions Term/Code not Available	A0908	Incorrect, Inadequate or Imprecise Result or Readings		
							E2403	No Clinical Signs, Symptoms or Conditions	A0908	Incorrect, Inadequate or Imprecise Result or Readings		
							F24	Insufficient Health Impact Information	A0908	Incorrect, Inadequate or Imprecise Result or Readings		
					CT1685	Cardiac pacemakers	E2402	Appropriate Clinical Signs, Symptoms and Conditions Term/Code not Available	A0908	Incorrect, Inadequate or Imprecise Result or Readings		
							E2403	No Clinical Signs, Symptoms or Conditions	A0908	Incorrect, Inadequate or Imprecise Result or Readings		
							F24	Insufficient Health Impact Information	A0908	Incorrect, Inadequate or Imprecise Result or Readings		
					CT233	Surgical	E2402	Appropriate Clinical Signs, Symptoms and Conditions Term/Code not Available	A0908	Incorrect, Inadequate or Imprecise Result or Readings		
							E2403	No Clinical Signs, Symptoms or Conditions	A0908	Incorrect, Inadequate or Imprecise Result or Readings		
							F24	Insufficient Health Impact Information	A0908	Incorrect, Inadequate or Imprecise Result or Readings		
					CT2406	Implantable	E2402	Appropriate Clinical Signs, Symptoms and Conditions Term/Code not Available	A0908	Incorrect, Inadequate or Imprecise Result or Readings		

							E2403	No Clinical Signs, Symptoms or Conditions	A0908	Incorrect, Inadequate or Imprecise Result or Readings		
							F24	Insufficient Health Impact Information	A0908	Incorrect, Inadequate or Imprecise Result or Readings		
					CT2407	Active implantable	E2402	Appropriate Clinical Signs, Symptoms and Conditions Term/Code not Available	A0908	Incorrect, Inadequate or Imprecise Result or Readings		
							E2403	No Clinical Signs, Symptoms or Conditions	A0908	Incorrect, Inadequate or Imprecise Result or Readings		
							F24	Insufficient Health Impact Information	A0908	Incorrect, Inadequate or Imprecise Result or Readings		
					CT257	Implantable cardiac pacemakers	E2402	Appropriate Clinical Signs, Symptoms and Conditions Term/Code not Available	A0908	Incorrect, Inadequate or Imprecise Result or Readings		
							E2403	No Clinical Signs, Symptoms or Conditions	A0908	Incorrect, Inadequate or Imprecise Result or Readings		
							F24	Insufficient Health Impact Information	A0908	Incorrect, Inadequate or Imprecise Result or Readings		
					CT2727	Active implant devices/ancillaries	E2402	Appropriate Clinical Signs, Symptoms and Conditions Term/Code not Available	A0908	Incorrect, Inadequate or Imprecise Result or Readings		
							E2403	No Clinical Signs, Symptoms or Conditions	A0908	Incorrect, Inadequate or Imprecise Result or Readings		
							F24	Insufficient Health Impact Information	A0908	Incorrect, Inadequate or Imprecise Result or Readings		
					CT2750	Implantable cardiac electrical stimulators and associated devices	E2402	Appropriate Clinical Signs, Symptoms and Conditions Term/Code not Available	A0908	Incorrect, Inadequate or Imprecise Result or Readings		
							E2403	No Clinical Signs, Symptoms or Conditions	A0908	Incorrect, Inadequate or Imprecise Result or Readings		
							F24	Insufficient Health Impact Information	A0908	Incorrect, Inadequate or Imprecise Result or Readings		
					CT2866	Implantable cardiac electrical stimulators	E2402	Appropriate Clinical Signs, Symptoms and Conditions Term/Code not Available	A0908	Incorrect, Inadequate or Imprecise Result or Readings		
							E2403	No Clinical Signs, Symptoms or Conditions	A0908	Incorrect, Inadequate or Imprecise Result or Readings		
							F24	Insufficient Health Impact Information	A0908	Incorrect, Inadequate or Imprecise Result or Readings		
					CT2985	Active	E2402	Appropriate Clinical Signs, Symptoms and Conditions Term/Code not Available	A0908	Incorrect, Inadequate or Imprecise Result or Readings		
							E2403	No Clinical Signs, Symptoms or Conditions	A0908	Incorrect, Inadequate or Imprecise Result or Readings		
							F24	Insufficient Health Impact Information	A0908	Incorrect, Inadequate or Imprecise Result or Readings		
					CT3001	Implantable metal	E2402	Appropriate Clinical Signs, Symptoms and Conditions Term/Code not Available	A0908	Incorrect, Inadequate or Imprecise Result or Readings		
							E2403	No Clinical Signs, Symptoms or Conditions	A0908	Incorrect, Inadequate or Imprecise Result or Readings		
							F24	Insufficient Health Impact Information	A0908	Incorrect, Inadequate or Imprecise Result or Readings		
					CT3002	Implantable synthetic polymer	E2402	Appropriate Clinical Signs, Symptoms and Conditions Term/Code not Available	A0908	Incorrect, Inadequate or Imprecise Result or Readings		
							E2403	No Clinical Signs, Symptoms or Conditions	A0908	Incorrect, Inadequate or Imprecise Result or Readings		
							F24	Insufficient Health Impact Information	A0908	Incorrect, Inadequate or Imprecise Result or Readings		
					CT3006	Implantable inorganic material	E2402	Appropriate Clinical Signs, Symptoms and Conditions Term/Code not Available	A0908	Incorrect, Inadequate or Imprecise Result or Readings		
							E2403	No Clinical Signs, Symptoms or Conditions	A0908	Incorrect, Inadequate or Imprecise Result or Readings		
							F24	Insufficient Health Impact Information	A0908	Incorrect, Inadequate or Imprecise Result or Readings		
					CT314	Electrically powered	E2402	Appropriate Clinical Signs, Symptoms and Conditions Term/Code not Available	A0908	Incorrect, Inadequate or Imprecise Result or Readings		
							E2403	No Clinical Signs, Symptoms or Conditions	A0908	Incorrect, Inadequate or Imprecise Result or Readings		
							F24	Insufficient Health Impact Information	A0908	Incorrect, Inadequate or Imprecise Result or Readings		
					CT321	Long-term surgical invasive	E2402	Appropriate Clinical Signs, Symptoms and Conditions Term/Code not Available	A0908	Incorrect, Inadequate or Imprecise Result or Readings		
							E2403	No Clinical Signs, Symptoms or Conditions	A0908	Incorrect, Inadequate or Imprecise Result or Readings		
							F24	Insufficient Health Impact Information	A0908	Incorrect, Inadequate or Imprecise Result or Readings		
					CT449	Generators and associated devices	E2402	Appropriate Clinical Signs, Symptoms and Conditions Term/Code not Available	A0908	Incorrect, Inadequate or Imprecise Result or Readings		
							E2403	No Clinical Signs, Symptoms or Conditions	A0908	Incorrect, Inadequate or Imprecise Result or Readings		
							F24	Insufficient Health Impact Information	A0908	Incorrect, Inadequate or Imprecise Result or Readings		
					CT981	Single use	E2402	Appropriate Clinical Signs, Symptoms and Conditions Term/Code not Available	A0908	Incorrect, Inadequate or Imprecise Result or Readings		
							E2403	No Clinical Signs, Symptoms or Conditions	A0908	Incorrect, Inadequate or Imprecise Result or Readings		
							F24	Insufficient Health Impact Information	A0908	Incorrect, Inadequate or Imprecise Result or Readings		
					CT983	Surgical invasive	E2402	Appropriate Clinical Signs, Symptoms and Conditions Term/Code not Available	A0908	Incorrect, Inadequate or Imprecise Result or Readings		
							E2403	No Clinical Signs, Symptoms or Conditions	A0908	Incorrect, Inadequate or Imprecise Result or Readings		
							F24	Insufficient Health Impact Information	A0908	Incorrect, Inadequate or Imprecise Result or Readings		
9891	2020	All other reportable incidents	37265	Dual-chamber implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A072201	High Impedance		
							F24	Insufficient Health Impact Information	A072201	High Impedance		
9892	2020	All other reportable incidents	37265	Dual-chamber implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A072201	High Impedance		
							F26	No Health Consequences or Impact	A072201	High Impedance		
9893	2020	All other reportable incidents	35852	Single-chamber implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	24/09/2020	24/09/2020
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	24/09/2020	24/09/2020
							F12	Serious Injury/ Illness/ Impairment	A070504	Premature Discharge of Battery	24/09/2020	24/09/2020
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	24/09/2020	24/09/2020
9894	2020	All other reportable incidents	35852	Single-chamber implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	23/09/2020	23/09/2020
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	23/09/2020	23/09/2020

							F1903	Device Explantation	A070504	Premature Discharge of Battery	23/09/2020	23/09/2020
9895	2020	All other reportable incidents	35852	Single-chamber implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A1107	Data Problem		
							F26	No Health Consequences or Impact	A1107	Data Problem		
9896	2020	All other reportable incidents	35852	Single-chamber implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	17/09/2020	17/09/2020
							F19	Surgical Intervention	A070504	Premature Discharge of Battery	17/09/2020	17/09/2020
9897	2020	All other reportable incidents	37265	Dual-chamber implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A0712	Pacing Problem		
							F26	No Health Consequences or Impact	A0712	Pacing Problem		
9898	2020	All other reportable incidents	35852	Single-chamber implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A0712	Pacing Problem		
							F26	No Health Consequences or Impact	A0712	Pacing Problem		
9899	2020	All other reportable incidents	47267	Single-chamber implantable pacemaker, rate-responsive	CT141	Electrodes/Leads and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	05/11/2020	05/11/2020
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	05/11/2020	05/11/2020
							F12	Serious Injury/ Illness/ Impairment	A070504	Premature Discharge of Battery	05/11/2020	05/11/2020
							F1903	Device Explantation	A070504	Premature Discharge of Battery	05/11/2020	05/11/2020
					CT145	Cardiovascular devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	05/11/2020	05/11/2020
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	05/11/2020	05/11/2020
							F12	Serious Injury/ Illness/ Impairment	A070504	Premature Discharge of Battery	05/11/2020	05/11/2020
							F1903	Device Explantation	A070504	Premature Discharge of Battery	05/11/2020	05/11/2020
					CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	05/11/2020	05/11/2020
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	05/11/2020	05/11/2020
							F12	Serious Injury/ Illness/ Impairment	A070504	Premature Discharge of Battery	05/11/2020	05/11/2020
							F1903	Device Explantation	A070504	Premature Discharge of Battery	05/11/2020	05/11/2020
					CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	05/11/2020	05/11/2020
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	05/11/2020	05/11/2020
							F12	Serious Injury/ Illness/ Impairment	A070504	Premature Discharge of Battery	05/11/2020	05/11/2020
							F1903	Device Explantation	A070504	Premature Discharge of Battery	05/11/2020	05/11/2020
					CT233	Surgical	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	05/11/2020	05/11/2020
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	05/11/2020	05/11/2020
							F12	Serious Injury/ Illness/ Impairment	A070504	Premature Discharge of Battery	05/11/2020	05/11/2020
							F1903	Device Explantation	A070504	Premature Discharge of Battery	05/11/2020	05/11/2020
					CT2406	Implantable	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	05/11/2020	05/11/2020
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	05/11/2020	05/11/2020
							F12	Serious Injury/ Illness/ Impairment	A070504	Premature Discharge of Battery	05/11/2020	05/11/2020
							F1903	Device Explantation	A070504	Premature Discharge of Battery	05/11/2020	05/11/2020
					CT2407	Active implantable	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	05/11/2020	05/11/2020
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	05/11/2020	05/11/2020
							F12	Serious Injury/ Illness/ Impairment	A070504	Premature Discharge of Battery	05/11/2020	05/11/2020
							F1903	Device Explantation	A070504	Premature Discharge of Battery	05/11/2020	05/11/2020
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	05/11/2020	05/11/2020
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	05/11/2020	05/11/2020
							F12	Serious Injury/ Illness/ Impairment	A070504	Premature Discharge of Battery	05/11/2020	05/11/2020
							F1903	Device Explantation	A070504	Premature Discharge of Battery	05/11/2020	05/11/2020
					CT2727	Active implant devices/ancillaries	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	05/11/2020	05/11/2020
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	05/11/2020	05/11/2020
							F12	Serious Injury/ Illness/ Impairment	A070504	Premature Discharge of Battery	05/11/2020	05/11/2020
							F1903	Device Explantation	A070504	Premature Discharge of Battery	05/11/2020	05/11/2020
					CT2750	Implantable cardiac electrical stimulators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	05/11/2020	05/11/2020
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	05/11/2020	05/11/2020
							F12	Serious Injury/ Illness/ Impairment	A070504	Premature Discharge of Battery	05/11/2020	05/11/2020
							F1903	Device Explantation	A070504	Premature Discharge of Battery	05/11/2020	05/11/2020
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	05/11/2020	05/11/2020
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	05/11/2020	05/11/2020
							F12	Serious Injury/ Illness/ Impairment	A070504	Premature Discharge of Battery	05/11/2020	05/11/2020
							F1903	Device Explantation	A070504	Premature Discharge of Battery	05/11/2020	05/11/2020
					CT2985	Active	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	05/11/2020	05/11/2020
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	05/11/2020	05/11/2020
							F12	Serious Injury/ Illness/ Impairment	A070504	Premature Discharge of Battery	05/11/2020	05/11/2020
							F1903	Device Explantation	A070504	Premature Discharge of Battery	05/11/2020	05/11/2020
					CT3001	Implantable metal	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	05/11/2020	05/11/2020
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	05/11/2020	05/11/2020
							F12	Serious Injury/ Illness/ Impairment	A070504	Premature Discharge of Battery	05/11/2020	05/11/2020
							F1903	Device Explantation	A070504	Premature Discharge of Battery	05/11/2020	05/11/2020
					CT3002	Implantable synthetic polymer	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	05/11/2020	05/11/2020
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	05/11/2020	05/11/2020
							F12	Serious Injury/ Illness/ Impairment	A070504	Premature Discharge of Battery	05/11/2020	05/11/2020
							F1903	Device Explantation	A070504	Premature Discharge of Battery	05/11/2020	05/11/2020
					CT3006	Implantable inorganic material	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	05/11/2020	05/11/2020
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	05/11/2020	05/11/2020
							F12	Serious Injury/ Illness/ Impairment	A070504	Premature Discharge of Battery	05/11/2020	05/11/2020
							F1903	Device Explantation	A070504	Premature Discharge of Battery	05/11/2020	05/11/2020
					CT314	Electrically powered	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	05/11/2020	05/11/2020
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	05/11/2020	05/11/2020
							F12	Serious Injury/ Illness/ Impairment	A070504	Premature Discharge of Battery	05/11/2020	05/11/2020
							F1903	Device Explantation	A070504	Premature Discharge of Battery	05/11/2020	05/11/2020
					CT321	Long-term surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	05/11/2020	05/11/2020
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	05/11/2020	05/11/2020
							F12	Serious Injury/ Illness/ Impairment	A070504	Premature Discharge of Battery	05/11/2020	05/11/2020
							F1903	Device Explantation	A070504	Premature Discharge of Battery	05/11/2020	05/11/2020
					CT449	Generators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	05/11/2020	05/11/2020
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	05/11/2020	05/11/2020
							F12	Serious Injury/ Illness/ Impairment	A070504	Premature Discharge of Battery	05/11/2020	05/11/2020
							F1903	Device Explantation	A070504	Premature Discharge of Battery	05/11/2020	05/11/2020
					CT981	Single use	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	05/11/2020	05/11/2020
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	05/11/2020	05/11/2020
							F12	Serious Injury/ Illness/ Impairment	A070504	Premature Discharge of Battery	05/11/2020	05/11/2020
							F1903	Device Explantation	A070504	Premature Discharge of Battery	05/11/2020	05/11/2020
					CT983	Surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	05/11/2020	05/11/2020
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	05/11/2020	05/11/2020
							F12	Serious Injury/ Illness/ Impairment	A070504	Premature Discharge of Battery	05/11/2020	05/11/2020

							F1903	Device Explantation	A070504	Premature Discharge of Battery	05/11/2020	05/11/2020
9900	2020	All other reportable incidents	35852	Single-chamber implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A070909	Oversensing		
							F26	No Health Consequences or Impact	A070909	Oversensing		
9901	2020	All other reportable incidents	35852	Single-chamber implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	26/08/2020	09/08/2016
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	26/08/2020	09/08/2016
9902	2020	All other reportable incidents	47270	Cardiac resynchronization therapy implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	10/09/2020	10/09/2020
							F1901	Additional Surgery	A070504	Premature Discharge of Battery	10/09/2020	10/09/2020
9903	2020	All other reportable incidents	60789	Intracardiac pacemaker	CT130	Catheters and associated devices	E0112	Dizziness	A0712	Pacing Problem		
							E011903	Syncope/Fainting	A0712	Pacing Problem		
							E2007	Fall	A0712	Pacing Problem		
							E2330	Pain	A0712	Pacing Problem		
							E233001	Chest Pain	A0712	Pacing Problem		
							F12	Serious Injury/ Illness/ Impairment	A0712	Pacing Problem		
					CT145	Cardiovascular devices	E0112	Dizziness	A0712	Pacing Problem		
							E011903	Syncope/Fainting	A0712	Pacing Problem		
							E2007	Fall	A0712	Pacing Problem		
							E2330	Pain	A0712	Pacing Problem		
							E233001	Chest Pain	A0712	Pacing Problem		
							F12	Serious Injury/ Illness/ Impairment	A0712	Pacing Problem		
					CT1684	Cardiac pacemakers and associated devices	E0112	Dizziness	A0712	Pacing Problem		
							E011903	Syncope/Fainting	A0712	Pacing Problem		
							E2007	Fall	A0712	Pacing Problem		
							E2330	Pain	A0712	Pacing Problem		
							E233001	Chest Pain	A0712	Pacing Problem		
							F12	Serious Injury/ Illness/ Impairment	A0712	Pacing Problem		
					CT1685	Cardiac pacemakers	E0112	Dizziness	A0712	Pacing Problem		
							E011903	Syncope/Fainting	A0712	Pacing Problem		
							E2007	Fall	A0712	Pacing Problem		
							E2330	Pain	A0712	Pacing Problem		
							E233001	Chest Pain	A0712	Pacing Problem		
							F12	Serious Injury/ Illness/ Impairment	A0712	Pacing Problem		
					CT233	Surgical	E0112	Dizziness	A0712	Pacing Problem		
							E011903	Syncope/Fainting	A0712	Pacing Problem		
							E2007	Fall	A0712	Pacing Problem		
							E2330	Pain	A0712	Pacing Problem		
							E233001	Chest Pain	A0712	Pacing Problem		
							F12	Serious Injury/ Illness/ Impairment	A0712	Pacing Problem		
					CT2406	Implantable	E0112	Dizziness	A0712	Pacing Problem		
							E011903	Syncope/Fainting	A0712	Pacing Problem		
							E2007	Fall	A0712	Pacing Problem		
							E2330	Pain	A0712	Pacing Problem		
							E233001	Chest Pain	A0712	Pacing Problem		
							F12	Serious Injury/ Illness/ Impairment	A0712	Pacing Problem		
					CT2407	Active implantable	E0112	Dizziness	A0712	Pacing Problem		
							E011903	Syncope/Fainting	A0712	Pacing Problem		
							E2007	Fall	A0712	Pacing Problem		
							E2330	Pain	A0712	Pacing Problem		
							E233001	Chest Pain	A0712	Pacing Problem		
							F12	Serious Injury/ Illness/ Impairment	A0712	Pacing Problem		
					CT257	Implantable cardiac pacemakers	E0112	Dizziness	A0712	Pacing Problem		
							E011903	Syncope/Fainting	A0712	Pacing Problem		
							E2007	Fall	A0712	Pacing Problem		
							E2330	Pain	A0712	Pacing Problem		
							E233001	Chest Pain	A0712	Pacing Problem		
							F12	Serious Injury/ Illness/ Impairment	A0712	Pacing Problem		
					CT2727	Active implant devices/ancillaries	E0112	Dizziness	A0712	Pacing Problem		
							E011903	Syncope/Fainting	A0712	Pacing Problem		
							E2007	Fall	A0712	Pacing Problem		
							E2330	Pain	A0712	Pacing Problem		
							E233001	Chest Pain	A0712	Pacing Problem		
							F12	Serious Injury/ Illness/ Impairment	A0712	Pacing Problem		
					CT2750	Implantable cardiac electrical stimulators and associated devices	E0112	Dizziness	A0712	Pacing Problem		
							E011903	Syncope/Fainting	A0712	Pacing Problem		
							E2007	Fall	A0712	Pacing Problem		
							E2330	Pain	A0712	Pacing Problem		
							E233001	Chest Pain	A0712	Pacing Problem		
							F12	Serious Injury/ Illness/ Impairment	A0712	Pacing Problem		
					CT2818	Catheterization/Cannulation/Fluid path devices	E0112	Dizziness	A0712	Pacing Problem		
							E011903	Syncope/Fainting	A0712	Pacing Problem		
							E2007	Fall	A0712	Pacing Problem		
							E2330	Pain	A0712	Pacing Problem		
							E233001	Chest Pain	A0712	Pacing Problem		
							F12	Serious Injury/ Illness/ Impairment	A0712	Pacing Problem		
					CT2866	Implantable cardiac electrical stimulators	E0112	Dizziness	A0712	Pacing Problem		
							E011903	Syncope/Fainting	A0712	Pacing Problem		
							E2007	Fall	A0712	Pacing Problem		
							E2330	Pain	A0712	Pacing Problem		
							E233001	Chest Pain	A0712	Pacing Problem		
							F12	Serious Injury/ Illness/ Impairment	A0712	Pacing Problem		
					CT2985	Active	E0112	Dizziness	A0712	Pacing Problem		
							E011903	Syncope/Fainting	A0712	Pacing Problem		
							E2007	Fall	A0712	Pacing Problem		
							E2330	Pain	A0712	Pacing Problem		
							E233001	Chest Pain	A0712	Pacing Problem		
							F12	Serious Injury/ Illness/ Impairment	A0712	Pacing Problem		
					CT3001	Implantable metal	E0112	Dizziness	A0712	Pacing Problem		
							E011903	Syncope/Fainting	A0712	Pacing Problem		
							E2007	Fall	A0712	Pacing Problem		

							E2330	Pain	A0712	Pacing Problem		
							E233001	Chest Pain	A0712	Pacing Problem		
							F12	Serious Injury/ Illness/ Impairment	A0712	Pacing Problem		
					CT3002	Implantable synthetic polymer	E0112	Dizziness	A0712	Pacing Problem		
							E011903	Syncope/Fainting	A0712	Pacing Problem		
							E2007	Fall	A0712	Pacing Problem		
							E2330	Pain	A0712	Pacing Problem		
							E233001	Chest Pain	A0712	Pacing Problem		
							F12	Serious Injury/ Illness/ Impairment	A0712	Pacing Problem		
					CT3006	Implantable inorganic material	E0112	Dizziness	A0712	Pacing Problem		
							E011903	Syncope/Fainting	A0712	Pacing Problem		
							E2007	Fall	A0712	Pacing Problem		
							E2330	Pain	A0712	Pacing Problem		
							E233001	Chest Pain	A0712	Pacing Problem		
							F12	Serious Injury/ Illness/ Impairment	A0712	Pacing Problem		
					CT314	Electrically powered	E0112	Dizziness	A0712	Pacing Problem		
							E011903	Syncope/Fainting	A0712	Pacing Problem		
							E2007	Fall	A0712	Pacing Problem		
							E2330	Pain	A0712	Pacing Problem		
							E233001	Chest Pain	A0712	Pacing Problem		
							F12	Serious Injury/ Illness/ Impairment	A0712	Pacing Problem		
					CT321	Long-term surgical invasive	E0112	Dizziness	A0712	Pacing Problem		
							E011903	Syncope/Fainting	A0712	Pacing Problem		
							E2007	Fall	A0712	Pacing Problem		
							E2330	Pain	A0712	Pacing Problem		
							E233001	Chest Pain	A0712	Pacing Problem		
							F12	Serious Injury/ Illness/ Impairment	A0712	Pacing Problem		
					CT981	Single use	E0112	Dizziness	A0712	Pacing Problem		
							E011903	Syncope/Fainting	A0712	Pacing Problem		
							E2007	Fall	A0712	Pacing Problem		
							E2330	Pain	A0712	Pacing Problem		
							E233001	Chest Pain	A0712	Pacing Problem		
							F12	Serious Injury/ Illness/ Impairment	A0712	Pacing Problem		
					CT983	Surgical invasive	E0112	Dizziness	A0712	Pacing Problem		
							E011903	Syncope/Fainting	A0712	Pacing Problem		
							E2007	Fall	A0712	Pacing Problem		
							E2330	Pain	A0712	Pacing Problem		
							E233001	Chest Pain	A0712	Pacing Problem		
							F12	Serious Injury/ Illness/ Impairment	A0712	Pacing Problem		
9904	2020	All other reportable incidents	47265	Dual-chamber implantable pacemaker, rate-responsive	CT141	Electrodes/Leads and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	30/09/2020	30/09/2020
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	30/09/2020	30/09/2020
							F12	Serious Injury/ Illness/ Impairment	A070504	Premature Discharge of Battery	30/09/2020	30/09/2020
							F1903	Device Explantation	A070504	Premature Discharge of Battery	30/09/2020	30/09/2020
					CT145	Cardiovascular devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	30/09/2020	30/09/2020
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	30/09/2020	30/09/2020
							F12	Serious Injury/ Illness/ Impairment	A070504	Premature Discharge of Battery	30/09/2020	30/09/2020
							F1903	Device Explantation	A070504	Premature Discharge of Battery	30/09/2020	30/09/2020
					CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	30/09/2020	30/09/2020
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	30/09/2020	30/09/2020
							F12	Serious Injury/ Illness/ Impairment	A070504	Premature Discharge of Battery	30/09/2020	30/09/2020
							F1903	Device Explantation	A070504	Premature Discharge of Battery	30/09/2020	30/09/2020
					CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	30/09/2020	30/09/2020
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	30/09/2020	30/09/2020
							F12	Serious Injury/ Illness/ Impairment	A070504	Premature Discharge of Battery	30/09/2020	30/09/2020
							F1903	Device Explantation	A070504	Premature Discharge of Battery	30/09/2020	30/09/2020
					CT233	Surgical	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	30/09/2020	30/09/2020
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	30/09/2020	30/09/2020
							F12	Serious Injury/ Illness/ Impairment	A070504	Premature Discharge of Battery	30/09/2020	30/09/2020
							F1903	Device Explantation	A070504	Premature Discharge of Battery	30/09/2020	30/09/2020
					CT2406	Implantable	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	30/09/2020	30/09/2020
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	30/09/2020	30/09/2020
							F12	Serious Injury/ Illness/ Impairment	A070504	Premature Discharge of Battery	30/09/2020	30/09/2020
							F1903	Device Explantation	A070504	Premature Discharge of Battery	30/09/2020	30/09/2020
					CT2407	Active implantable	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	30/09/2020	30/09/2020
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	30/09/2020	30/09/2020
							F12	Serious Injury/ Illness/ Impairment	A070504	Premature Discharge of Battery	30/09/2020	30/09/2020
							F1903	Device Explantation	A070504	Premature Discharge of Battery	30/09/2020	30/09/2020
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	30/09/2020	30/09/2020
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	30/09/2020	30/09/2020
							F12	Serious Injury/ Illness/ Impairment	A070504	Premature Discharge of Battery	30/09/2020	30/09/2020
							F1903	Device Explantation	A070504	Premature Discharge of Battery	30/09/2020	30/09/2020
					CT2727	Active implant devices/ancillaries	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	30/09/2020	30/09/2020
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	30/09/2020	30/09/2020
							F12	Serious Injury/ Illness/ Impairment	A070504	Premature Discharge of Battery	30/09/2020	30/09/2020
							F1903	Device Explantation	A070504	Premature Discharge of Battery	30/09/2020	30/09/2020
					CT2750	Implantable cardiac electrical stimulators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	30/09/2020	30/09/2020
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	30/09/2020	30/09/2020
							F12	Serious Injury/ Illness/ Impairment	A070504	Premature Discharge of Battery	30/09/2020	30/09/2020
							F1903	Device Explantation	A070504	Premature Discharge of Battery	30/09/2020	30/09/2020
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	30/09/2020	30/09/2020
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	30/09/2020	30/09/2020
							F12	Serious Injury/ Illness/ Impairment	A070504	Premature Discharge of Battery	30/09/2020	30/09/2020
							F1903	Device Explantation	A070504	Premature Discharge of Battery	30/09/2020	30/09/2020
					CT2985	Active	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	30/09/2020	30/09/2020
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	30/09/2020	30/09/2020
							F12	Serious Injury/ Illness/ Impairment	A070504	Premature Discharge of Battery	30/09/2020	30/09/2020
							F1903	Device Explantation	A070504	Premature Discharge of Battery	30/09/2020	30/09/2020
					CT3001	Implantable metal	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	30/09/2020	30/09/2020

							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	30/09/2020	30/09/2020
							F12	Serious Injury/ Illness/ Impairment	A070504	Premature Discharge of Battery	30/09/2020	30/09/2020
							F1903	Device Explantation	A070504	Premature Discharge of Battery	30/09/2020	30/09/2020
					CT3002	Implantable synthetic polymer	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	30/09/2020	30/09/2020
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	30/09/2020	30/09/2020
							F12	Serious Injury/ Illness/ Impairment	A070504	Premature Discharge of Battery	30/09/2020	30/09/2020
							F1903	Device Explantation	A070504	Premature Discharge of Battery	30/09/2020	30/09/2020
					CT3006	Implantable inorganic material	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	30/09/2020	30/09/2020
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	30/09/2020	30/09/2020
							F12	Serious Injury/ Illness/ Impairment	A070504	Premature Discharge of Battery	30/09/2020	30/09/2020
							F1903	Device Explantation	A070504	Premature Discharge of Battery	30/09/2020	30/09/2020
					CT314	Electrically powered	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	30/09/2020	30/09/2020
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	30/09/2020	30/09/2020
							F12	Serious Injury/ Illness/ Impairment	A070504	Premature Discharge of Battery	30/09/2020	30/09/2020
							F1903	Device Explantation	A070504	Premature Discharge of Battery	30/09/2020	30/09/2020
					CT321	Long-term surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	30/09/2020	30/09/2020
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	30/09/2020	30/09/2020
							F12	Serious Injury/ Illness/ Impairment	A070504	Premature Discharge of Battery	30/09/2020	30/09/2020
							F1903	Device Explantation	A070504	Premature Discharge of Battery	30/09/2020	30/09/2020
					CT449	Generators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	30/09/2020	30/09/2020
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	30/09/2020	30/09/2020
							F12	Serious Injury/ Illness/ Impairment	A070504	Premature Discharge of Battery	30/09/2020	30/09/2020
							F1903	Device Explantation	A070504	Premature Discharge of Battery	30/09/2020	30/09/2020
					CT981	Single use	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	30/09/2020	30/09/2020
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	30/09/2020	30/09/2020
							F12	Serious Injury/ Illness/ Impairment	A070504	Premature Discharge of Battery	30/09/2020	30/09/2020
							F1903	Device Explantation	A070504	Premature Discharge of Battery	30/09/2020	30/09/2020
					CT983	Surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	30/09/2020	30/09/2020
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	30/09/2020	30/09/2020
							F12	Serious Injury/ Illness/ Impairment	A070504	Premature Discharge of Battery	30/09/2020	30/09/2020
							F1903	Device Explantation	A070504	Premature Discharge of Battery	30/09/2020	30/09/2020
9905	2019	All other reportable incidents	61438	Cardiac pulse generator reader, home-use	CT1684	Cardiac pacemakers and associated devices	E1704	Burn(s)	A1005	Overheating of Device		
9906	2019	All other reportable incidents	35223	Endocardial/interventricular septal pacing lead	CT1684	Cardiac pacemakers and associated devices	E233001	Chest Pain	A010402	Migration	18/11/2018	
							F1905	Device Revision or Replacement	A010402	Migration	18/11/2018	
9907	2019	All other reportable incidents	36241	Pacemaker/defibrillator lead	CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A0722	Impedance Problem	13/09/2019	13/09/2019
9908	2019	All other reportable incidents	17882	Physiologic-monitoring defibrillation system	CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A071301	Failure to Deliver Shock/Stimulation		
9909	2019	All other reportable incidents	36241	Pacemaker/defibrillator lead	CT1684	Cardiac pacemakers and associated devices	E0506	Hemorrhage/Blood Loss/Bleeding	A0104	Migration or Expulsion of Device	20/09/2019	
							F1905	Device Revision or Replacement	A0104	Migration or Expulsion of Device	20/09/2019	
9910	2019	All other reportable incidents	47265	Dual-chamber implantable pacemaker, rate-responsive	CT141	Electrodes/Leads and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A	Annex A		
					CT145	Cardiovascular devices	E2403	No Clinical Signs, Symptoms or Conditions	A	Annex A		
					CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A	Annex A		
					CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A	Annex A		
					CT233	Surgical	E2403	No Clinical Signs, Symptoms or Conditions	A	Annex A		
					CT2406	Implantable	E2403	No Clinical Signs, Symptoms or Conditions	A	Annex A		
					CT2407	Active implantable	E2403	No Clinical Signs, Symptoms or Conditions	A	Annex A		
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A	Annex A		
					CT2727	Active implant devices/ancillaries	E2403	No Clinical Signs, Symptoms or Conditions	A	Annex A		
					CT2750	Implantable cardiac electrical stimulators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A	Annex A		
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A	Annex A		
					CT2985	Active	E2403	No Clinical Signs, Symptoms or Conditions	A	Annex A		
					CT3001	Implantable metal	E2403	No Clinical Signs, Symptoms or Conditions	A	Annex A		
					CT3002	Implantable synthetic polymer	E2403	No Clinical Signs, Symptoms or Conditions	A	Annex A		
					CT3006	Implantable inorganic material	E2403	No Clinical Signs, Symptoms or Conditions	A	Annex A		
					CT314	Electrically powered	E2403	No Clinical Signs, Symptoms or Conditions	A	Annex A		
					CT321	Long-term surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A	Annex A		
					CT449	Generators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A	Annex A		
					CT981	Single use	E2403	No Clinical Signs, Symptoms or Conditions	A	Annex A		
					CT983	Surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A	Annex A		
9911	2020	All other reportable incidents	47270	Cardiac resynchronization therapy implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F1901	Additional Surgery	A070504	Premature Discharge of Battery		
9912	2020	All other reportable incidents	47265	Dual-chamber implantable pacemaker, rate-responsive	CT141	Electrodes/Leads and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070101	Failure to Capture	15/07/2020	15/07/2020
							F1901	Additional Surgery	A070101	Failure to Capture	15/07/2020	15/07/2020
							F1905	Device Revision or Replacement	A070101	Failure to Capture	15/07/2020	15/07/2020
					CT145	Cardiovascular devices	E2403	No Clinical Signs, Symptoms or Conditions	A070101	Failure to Capture	15/07/2020	15/07/2020
							F1901	Additional Surgery	A070101	Failure to Capture	15/07/2020	15/07/2020
							F1905	Device Revision or Replacement	A070101	Failure to Capture	15/07/2020	15/07/2020
					CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070101	Failure to Capture	15/07/2020	15/07/2020
							F1901	Additional Surgery	A070101	Failure to Capture	15/07/2020	15/07/2020
							F1905	Device Revision or Replacement	A070101	Failure to Capture	15/07/2020	15/07/2020
					CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A070101	Failure to Capture	15/07/2020	15/07/2020
							F1901	Additional Surgery	A070101	Failure to Capture	15/07/2020	15/07/2020
							F1905	Device Revision or Replacement	A070101	Failure to Capture	15/07/2020	15/07/2020
					CT233	Surgical	E2403	No Clinical Signs, Symptoms or Conditions	A070101	Failure to Capture	15/07/2020	15/07/2020
							F1901	Additional Surgery	A070101	Failure to Capture	15/07/2020	15/07/2020
							F1905	Device Revision or Replacement	A070101	Failure to Capture	15/07/2020	15/07/2020
					CT2406	Implantable	E2403	No Clinical Signs, Symptoms or Conditions	A070101	Failure to Capture	15/07/2020	15/07/2020
							F1901	Additional Surgery	A070101	Failure to Capture	15/07/2020	15/07/2020
							F1905	Device Revision or Replacement	A070101	Failure to Capture	15/07/2020	15/07/2020
					CT2407	Active implantable	E2403	No Clinical Signs, Symptoms or Conditions	A070101	Failure to Capture	15/07/2020	15/07/2020
							F1901	Additional Surgery	A070101	Failure to Capture	15/07/2020	15/07/2020
							F1905	Device Revision or Replacement	A070101	Failure to Capture	15/07/2020	15/07/2020
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A070101	Failure to Capture	15/07/2020	15/07/2020
							F1901	Additional Surgery	A070101	Failure to Capture	15/07/2020	15/07/2020
							F1905	Device Revision or Replacement	A070101	Failure to Capture	15/07/2020	15/07/2020
					CT2727	Active implant devices/ancillaries	E2403	No Clinical Signs, Symptoms or Conditions	A070101	Failure to Capture	15/07/2020	15/07/2020
							F1901	Additional Surgery	A070101	Failure to Capture	15/07/2020	15/07/2020
							F1905	Device Revision or Replacement	A070101	Failure to Capture	15/07/2020	15/07/2020

					CT2750	Implantable cardiac electrical stimulators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070101	Failure to Capture	15/07/2020	15/07/2020
							F1901	Additional Surgery	A070101	Failure to Capture	15/07/2020	15/07/2020
							F1905	Device Revision or Replacement	A070101	Failure to Capture	15/07/2020	15/07/2020
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A070101	Failure to Capture	15/07/2020	15/07/2020
							F1901	Additional Surgery	A070101	Failure to Capture	15/07/2020	15/07/2020
							F1905	Device Revision or Replacement	A070101	Failure to Capture	15/07/2020	15/07/2020
					CT2985	Active	E2403	No Clinical Signs, Symptoms or Conditions	A070101	Failure to Capture	15/07/2020	15/07/2020
							F1901	Additional Surgery	A070101	Failure to Capture	15/07/2020	15/07/2020
							F1905	Device Revision or Replacement	A070101	Failure to Capture	15/07/2020	15/07/2020
					CT3001	Implantable metal	E2403	No Clinical Signs, Symptoms or Conditions	A070101	Failure to Capture	15/07/2020	15/07/2020
							F1901	Additional Surgery	A070101	Failure to Capture	15/07/2020	15/07/2020
							F1905	Device Revision or Replacement	A070101	Failure to Capture	15/07/2020	15/07/2020
					CT3002	Implantable synthetic polymer	E2403	No Clinical Signs, Symptoms or Conditions	A070101	Failure to Capture	15/07/2020	15/07/2020
							F1901	Additional Surgery	A070101	Failure to Capture	15/07/2020	15/07/2020
							F1905	Device Revision or Replacement	A070101	Failure to Capture	15/07/2020	15/07/2020
					CT3006	Implantable inorganic material	E2403	No Clinical Signs, Symptoms or Conditions	A070101	Failure to Capture	15/07/2020	15/07/2020
							F1901	Additional Surgery	A070101	Failure to Capture	15/07/2020	15/07/2020
							F1905	Device Revision or Replacement	A070101	Failure to Capture	15/07/2020	15/07/2020
					CT314	Electrically powered	E2403	No Clinical Signs, Symptoms or Conditions	A070101	Failure to Capture	15/07/2020	15/07/2020
							F1901	Additional Surgery	A070101	Failure to Capture	15/07/2020	15/07/2020
							F1905	Device Revision or Replacement	A070101	Failure to Capture	15/07/2020	15/07/2020
					CT321	Long-term surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A070101	Failure to Capture	15/07/2020	15/07/2020
							F1901	Additional Surgery	A070101	Failure to Capture	15/07/2020	15/07/2020
							F1905	Device Revision or Replacement	A070101	Failure to Capture	15/07/2020	15/07/2020
					CT449	Generators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070101	Failure to Capture	15/07/2020	15/07/2020
							F1901	Additional Surgery	A070101	Failure to Capture	15/07/2020	15/07/2020
							F1905	Device Revision or Replacement	A070101	Failure to Capture	15/07/2020	15/07/2020
					CT981	Single use	E2403	No Clinical Signs, Symptoms or Conditions	A070101	Failure to Capture	15/07/2020	15/07/2020
							F1901	Additional Surgery	A070101	Failure to Capture	15/07/2020	15/07/2020
							F1905	Device Revision or Replacement	A070101	Failure to Capture	15/07/2020	15/07/2020
					CT983	Surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A070101	Failure to Capture	15/07/2020	15/07/2020
							F1901	Additional Surgery	A070101	Failure to Capture	15/07/2020	15/07/2020
							F1905	Device Revision or Replacement	A070101	Failure to Capture	15/07/2020	15/07/2020
9913	2020	All other reportable incidents	60789	Intracardiac pacemaker	CT130	Catheters and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
									A071102	Failure to Interrogate		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
									A071102	Failure to Interrogate		
					CT145	Cardiovascular devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
									A071102	Failure to Interrogate		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
									A071102	Failure to Interrogate		
					CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
									A071102	Failure to Interrogate		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
									A071102	Failure to Interrogate		
					CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
									A071102	Failure to Interrogate		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
									A071102	Failure to Interrogate		
					CT233	Surgical	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
									A071102	Failure to Interrogate		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
									A071102	Failure to Interrogate		
					CT2406	Implantable	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
									A071102	Failure to Interrogate		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
									A071102	Failure to Interrogate		
					CT2407	Active implantable	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
									A071102	Failure to Interrogate		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
									A071102	Failure to Interrogate		
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
									A071102	Failure to Interrogate		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
									A071102	Failure to Interrogate		
					CT2727	Active implant devices/ancillaries	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
									A071102	Failure to Interrogate		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
									A071102	Failure to Interrogate		
					CT2750	Implantable cardiac electrical stimulators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
									A071102	Failure to Interrogate		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
									A071102	Failure to Interrogate		
					CT2818	Catheterization/Cannulation/Fluid path devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
									A071102	Failure to Interrogate		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
									A071102	Failure to Interrogate		
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
									A071102	Failure to Interrogate		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
									A071102	Failure to Interrogate		
					CT2985	Active	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
									A071102	Failure to Interrogate		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
									A071102	Failure to Interrogate		
					CT3001	Implantable metal	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
									A071102	Failure to Interrogate		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		

									A071102	Failure to Interrogate		
					CT3002	Implantable synthetic polymer	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
									A071102	Failure to Interrogate		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
									A071102	Failure to Interrogate		
					CT3006	Implantable inorganic material	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
									A071102	Failure to Interrogate		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
									A071102	Failure to Interrogate		
					CT314	Electrically powered	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
									A071102	Failure to Interrogate		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
									A071102	Failure to Interrogate		
					CT321	Long-term surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
									A071102	Failure to Interrogate		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
									A071102	Failure to Interrogate		
					CT981	Single use	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
									A071102	Failure to Interrogate		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
									A071102	Failure to Interrogate		
					CT983	Surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
									A071102	Failure to Interrogate		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
									A071102	Failure to Interrogate		
9914	2019	All other reportable incidents	35852	Single-chamber implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A070910	Undersensing		
9915	2020	All other reportable incidents	37265	Dual-chamber implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A070603	Failure to Charge	08/01/2021	08/01/2021
							F1901	Additional Surgery	A070603	Failure to Charge	08/01/2021	08/01/2021
9916	2019	All other reportable incidents	35852	Single-chamber implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A070909	Oversensing		
9917	2019	All other reportable incidents	47270	Cardiac resynchronization therapy implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A	Annex A		
9918	2020	All other reportable incidents	47265	Dual-chamber implantable pacemaker, rate-responsive	CT141	Electrodes/Leads and associated devices	E0717	Dyspnea	A090804	False Positive Result		
							E2312	Fatigue	A090804	False Positive Result		
							F15	Recognized Device or Procedural Complication	A090804	False Positive Result		
					CT145	Cardiovascular devices	E0717	Dyspnea	A090804	False Positive Result		
							E2312	Fatigue	A090804	False Positive Result		
							F15	Recognized Device or Procedural Complication	A090804	False Positive Result		
					CT1684	Cardiac pacemakers and associated devices	E0717	Dyspnea	A090804	False Positive Result		
							E2312	Fatigue	A090804	False Positive Result		
							F15	Recognized Device or Procedural Complication	A090804	False Positive Result		
					CT1685	Cardiac pacemakers	E0717	Dyspnea	A090804	False Positive Result		
							E2312	Fatigue	A090804	False Positive Result		
							F15	Recognized Device or Procedural Complication	A090804	False Positive Result		
					CT233	Surgical	E0717	Dyspnea	A090804	False Positive Result		
							E2312	Fatigue	A090804	False Positive Result		
							F15	Recognized Device or Procedural Complication	A090804	False Positive Result		
					CT2406	Implantable	E0717	Dyspnea	A090804	False Positive Result		
							E2312	Fatigue	A090804	False Positive Result		
							F15	Recognized Device or Procedural Complication	A090804	False Positive Result		
					CT2407	Active implantable	E0717	Dyspnea	A090804	False Positive Result		
							E2312	Fatigue	A090804	False Positive Result		
							F15	Recognized Device or Procedural Complication	A090804	False Positive Result		
					CT257	Implantable cardiac pacemakers	E0717	Dyspnea	A090804	False Positive Result		
							E2312	Fatigue	A090804	False Positive Result		
							F15	Recognized Device or Procedural Complication	A090804	False Positive Result		
					CT2727	Active implant devices/ancillaries	E0717	Dyspnea	A090804	False Positive Result		
							E2312	Fatigue	A090804	False Positive Result		
							F15	Recognized Device or Procedural Complication	A090804	False Positive Result		
					CT2750	Implantable cardiac electrical stimulators and associated devices	E0717	Dyspnea	A090804	False Positive Result		
							E2312	Fatigue	A090804	False Positive Result		
							F15	Recognized Device or Procedural Complication	A090804	False Positive Result		
					CT2866	Implantable cardiac electrical stimulators	E0717	Dyspnea	A090804	False Positive Result		
							E2312	Fatigue	A090804	False Positive Result		
							F15	Recognized Device or Procedural Complication	A090804	False Positive Result		
					CT2985	Active	E0717	Dyspnea	A090804	False Positive Result		
							E2312	Fatigue	A090804	False Positive Result		
							F15	Recognized Device or Procedural Complication	A090804	False Positive Result		
					CT3001	Implantable metal	E0717	Dyspnea	A090804	False Positive Result		
							E2312	Fatigue	A090804	False Positive Result		
							F15	Recognized Device or Procedural Complication	A090804	False Positive Result		
					CT3002	Implantable synthetic polymer	E0717	Dyspnea	A090804	False Positive Result		
							E2312	Fatigue	A090804	False Positive Result		
							F15	Recognized Device or Procedural Complication	A090804	False Positive Result		
					CT3006	Implantable inorganic material	E0717	Dyspnea	A090804	False Positive Result		
							E2312	Fatigue	A090804	False Positive Result		
							F15	Recognized Device or Procedural Complication	A090804	False Positive Result		
					CT314	Electrically powered	E0717	Dyspnea	A090804	False Positive Result		
							E2312	Fatigue	A090804	False Positive Result		
							F15	Recognized Device or Procedural Complication	A090804	False Positive Result		
					CT321	Long-term surgical invasive	E0717	Dyspnea	A090804	False Positive Result		
							E2312	Fatigue	A090804	False Positive Result		
							F15	Recognized Device or Procedural Complication	A090804	False Positive Result		
					CT449	Generators and associated devices	E0717	Dyspnea	A090804	False Positive Result		
							E2312	Fatigue	A090804	False Positive Result		
							F15	Recognized Device or Procedural Complication	A090804	False Positive Result		
					CT981	Single use	E0717	Dyspnea	A090804	False Positive Result		
							E2312	Fatigue	A090804	False Positive Result		
							F15	Recognized Device or Procedural Complication	A090804	False Positive Result		

					CT983	Surgical invasive	E0717	Dyspnea	A090804	False Positive Result			
							E2312	Fatigue	A090804	False Positive Result			
							F15	Recognized Device or Procedural Complication	A090804	False Positive Result			
9919	2019	All other reportable incidents	47265	Dual-chamber implantable pacemaker, rate-responsive	CT141	Electrodes/Leads and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	23/10/2019	23/10/2019	
					CT145	Cardiovascular devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	23/10/2019	23/10/2019	
					CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	23/10/2019	23/10/2019	
					CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	23/10/2019	23/10/2019	
					CT233	Surgical	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	23/10/2019	23/10/2019	
					CT2406	Implantable	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	23/10/2019	23/10/2019	
					CT2407	Active implantable	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	23/10/2019	23/10/2019	
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	23/10/2019	23/10/2019	
					CT2727	Active implant devices/ancillaries	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	23/10/2019	23/10/2019	
					CT2750	Implantable cardiac electrical stimulators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	23/10/2019	23/10/2019	
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	23/10/2019	23/10/2019	
					CT2985	Active	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	23/10/2019	23/10/2019	
					CT3001	Implantable metal	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	23/10/2019	23/10/2019	
					CT3002	Implantable synthetic polymer	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	23/10/2019	23/10/2019	
					CT3006	Implantable inorganic material	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	23/10/2019	23/10/2019	
					CT314	Electrically powered	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	23/10/2019	23/10/2019	
					CT321	Long-term surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	23/10/2019	23/10/2019	
					CT449	Generators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	23/10/2019	23/10/2019	
					CT981	Single use	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	23/10/2019	23/10/2019	
					CT983	Surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	23/10/2019	23/10/2019	
9920	2019	All other reportable incidents	47267	Single-chamber implantable pacemaker, rate-responsive	CT141	Electrodes/Leads and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery			
					CT145	Cardiovascular devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery			
					CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery			
					CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery			
					CT233	Surgical	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery			
					CT2406	Implantable	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery			
					CT2407	Active implantable	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery			
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery			
					CT2727	Active implant devices/ancillaries	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery			
					CT2750	Implantable cardiac electrical stimulators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery			
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery			
					CT2985	Active	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery			
					CT3001	Implantable metal	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery			
					CT3002	Implantable synthetic polymer	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery			
					CT3006	Implantable inorganic material	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery			
					CT314	Electrically powered	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery			
					CT321	Long-term surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery			
					CT449	Generators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery			
					CT981	Single use	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery			
					CT983	Surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery			
9921	2019	All other reportable incidents	35852	Single-chamber implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	F1905	Device Revision or Replacement	A070909	Oversensing			
9922	2020	All other reportable incidents	47265	Dual-chamber implantable pacemaker, rate-responsive	CT141	Electrodes/Leads and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070909	Oversensing			
							F26	No Health Consequences or Impact	A070909	Oversensing			
					CT145	Cardiovascular devices	E2403	No Clinical Signs, Symptoms or Conditions	A070909	Oversensing			
							F26	No Health Consequences or Impact	A070909	Oversensing			
					CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070909	Oversensing			
							F26	No Health Consequences or Impact	A070909	Oversensing			
					CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A070909	Oversensing			
							F26	No Health Consequences or Impact	A070909	Oversensing			
					CT233	Surgical	E2403	No Clinical Signs, Symptoms or Conditions	A070909	Oversensing			
							F26	No Health Consequences or Impact	A070909	Oversensing			
					CT2406	Implantable	E2403	No Clinical Signs, Symptoms or Conditions	A070909	Oversensing			
							F26	No Health Consequences or Impact	A070909	Oversensing			
					CT2407	Active implantable	E2403	No Clinical Signs, Symptoms or Conditions	A070909	Oversensing			
							F26	No Health Consequences or Impact	A070909	Oversensing			
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A070909	Oversensing			
							F26	No Health Consequences or Impact	A070909	Oversensing			
					CT2727	Active implant devices/ancillaries	E2403	No Clinical Signs, Symptoms or Conditions	A070909	Oversensing			
							F26	No Health Consequences or Impact	A070909	Oversensing			
					CT2750	Implantable cardiac electrical stimulators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070909	Oversensing			
							F26	No Health Consequences or Impact	A070909	Oversensing			
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A070909	Oversensing			
							F26	No Health Consequences or Impact	A070909	Oversensing			
					CT2985	Active	E2403	No Clinical Signs, Symptoms or Conditions	A070909	Oversensing			
							F26	No Health Consequences or Impact	A070909	Oversensing			
					CT3001	Implantable metal	E2403	No Clinical Signs, Symptoms or Conditions	A070909	Oversensing			
							F26	No Health Consequences or Impact	A070909	Oversensing			
					CT3002	Implantable synthetic polymer	E2403	No Clinical Signs, Symptoms or Conditions	A070909	Oversensing			
							F26	No Health Consequences or Impact	A070909	Oversensing			
					CT3006	Implantable inorganic material	E2403	No Clinical Signs, Symptoms or Conditions	A070909	Oversensing			
							F26	No Health Consequences or Impact	A070909	Oversensing			
					CT314	Electrically powered	E2403	No Clinical Signs, Symptoms or Conditions	A070909	Oversensing			
							F26	No Health Consequences or Impact	A070909	Oversensing			
					CT321	Long-term surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A070909	Oversensing			
							F26	No Health Consequences or Impact	A070909	Oversensing			
					CT449	Generators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070909	Oversensing			
							F26	No Health Consequences or Impact	A070909	Oversensing			
					CT981	Single use	E2403	No Clinical Signs, Symptoms or Conditions	A070909	Oversensing			
							F26	No Health Consequences or Impact	A070909	Oversensing			
					CT983	Surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A070909	Oversensing			
							F26	No Health Consequences or Impact	A070909	Oversensing			
9923	2019	All other reportable incidents	47265	Dual-chamber implantable pacemaker, rate-responsive	CT141	Electrodes/Leads and associated devices	F1905	Device Revision or Replacement	A071102	Failure to Interrogate	02/10/2019		
					CT145	Cardiovascular devices	F1905	Device Revision or Replacement	A071102	Failure to Interrogate	02/10/2019		

					CT1684	Cardiac pacemakers and associated devices	F1905	Device Revision or Replacement	A071102	Failure to Interrogate	02/10/2019		
					CT1685	Cardiac pacemakers	F1905	Device Revision or Replacement	A071102	Failure to Interrogate	02/10/2019		
					CT233	Surgical	F1905	Device Revision or Replacement	A071102	Failure to Interrogate	02/10/2019		
					CT2406	Implantable	F1905	Device Revision or Replacement	A071102	Failure to Interrogate	02/10/2019		
					CT2407	Active implantable	F1905	Device Revision or Replacement	A071102	Failure to Interrogate	02/10/2019		
					CT257	Implantable cardiac pacemakers	F1905	Device Revision or Replacement	A071102	Failure to Interrogate	02/10/2019		
					CT2727	Active implant devices/ancillaries	F1905	Device Revision or Replacement	A071102	Failure to Interrogate	02/10/2019		
					CT2750	Implantable cardiac electrical stimulators and associated devices	F1905	Device Revision or Replacement	A071102	Failure to Interrogate	02/10/2019		
					CT2866	Implantable cardiac electrical stimulators	F1905	Device Revision or Replacement	A071102	Failure to Interrogate	02/10/2019		
					CT2985	Active	F1905	Device Revision or Replacement	A071102	Failure to Interrogate	02/10/2019		
					CT3001	Implantable metal	F1905	Device Revision or Replacement	A071102	Failure to Interrogate	02/10/2019		
					CT3002	Implantable synthetic polymer	F1905	Device Revision or Replacement	A071102	Failure to Interrogate	02/10/2019		
					CT3006	Implantable inorganic material	F1905	Device Revision or Replacement	A071102	Failure to Interrogate	02/10/2019		
					CT314	Electrically powered	F1905	Device Revision or Replacement	A071102	Failure to Interrogate	02/10/2019		
					CT321	Long-term surgical invasive	F1905	Device Revision or Replacement	A071102	Failure to Interrogate	02/10/2019		
					CT449	Generators and associated devices	F1905	Device Revision or Replacement	A071102	Failure to Interrogate	02/10/2019		
					CT981	Single use	F1905	Device Revision or Replacement	A071102	Failure to Interrogate	02/10/2019		
					CT983	Surgical invasive	F1905	Device Revision or Replacement	A071102	Failure to Interrogate	02/10/2019		
9924	2020	All other reportable incidents	60789	Intracardiac pacemaker	CT130	Catheters and associated devices	E011903	Syncope/Fainting	A24	No Apparent Device Problem			
							E0602	Cardiac Arrest	A24	No Apparent Device Problem			
							E0604	Cardiac Perforation	A24	No Apparent Device Problem			
							E0605	Cardiac Tamponade	A24	No Apparent Device Problem			
							F08	Hospitalization or Prolonged Hospitalization	A24	No Apparent Device Problem			
							F1203	Life Threatening Illness or Injury	A24	No Apparent Device Problem			
					CT145	Cardiovascular devices	E011903	Syncope/Fainting	A24	No Apparent Device Problem			
							E0602	Cardiac Arrest	A24	No Apparent Device Problem			
							E0604	Cardiac Perforation	A24	No Apparent Device Problem			
							E0605	Cardiac Tamponade	A24	No Apparent Device Problem			
							F08	Hospitalization or Prolonged Hospitalization	A24	No Apparent Device Problem			
							F1203	Life Threatening Illness or Injury	A24	No Apparent Device Problem			
					CT1684	Cardiac pacemakers and associated devices	E011903	Syncope/Fainting	A24	No Apparent Device Problem			
							E0602	Cardiac Arrest	A24	No Apparent Device Problem			
							E0604	Cardiac Perforation	A24	No Apparent Device Problem			
							E0605	Cardiac Tamponade	A24	No Apparent Device Problem			
							F08	Hospitalization or Prolonged Hospitalization	A24	No Apparent Device Problem			
							F1203	Life Threatening Illness or Injury	A24	No Apparent Device Problem			
					CT1685	Cardiac pacemakers	E011903	Syncope/Fainting	A24	No Apparent Device Problem			
							E0602	Cardiac Arrest	A24	No Apparent Device Problem			
							E0604	Cardiac Perforation	A24	No Apparent Device Problem			
							E0605	Cardiac Tamponade	A24	No Apparent Device Problem			
							F08	Hospitalization or Prolonged Hospitalization	A24	No Apparent Device Problem			
							F1203	Life Threatening Illness or Injury	A24	No Apparent Device Problem			
					CT233	Surgical	E011903	Syncope/Fainting	A24	No Apparent Device Problem			
							E0602	Cardiac Arrest	A24	No Apparent Device Problem			
							E0604	Cardiac Perforation	A24	No Apparent Device Problem			
							E0605	Cardiac Tamponade	A24	No Apparent Device Problem			
							F08	Hospitalization or Prolonged Hospitalization	A24	No Apparent Device Problem			
							F1203	Life Threatening Illness or Injury	A24	No Apparent Device Problem			
					CT2406	Implantable	E011903	Syncope/Fainting	A24	No Apparent Device Problem			
							E0602	Cardiac Arrest	A24	No Apparent Device Problem			
							E0604	Cardiac Perforation	A24	No Apparent Device Problem			
							E0605	Cardiac Tamponade	A24	No Apparent Device Problem			
							F08	Hospitalization or Prolonged Hospitalization	A24	No Apparent Device Problem			
							F1203	Life Threatening Illness or Injury	A24	No Apparent Device Problem			
					CT2407	Active implantable	E011903	Syncope/Fainting	A24	No Apparent Device Problem			
							E0602	Cardiac Arrest	A24	No Apparent Device Problem			
							E0604	Cardiac Perforation	A24	No Apparent Device Problem			
							E0605	Cardiac Tamponade	A24	No Apparent Device Problem			
							F08	Hospitalization or Prolonged Hospitalization	A24	No Apparent Device Problem			
							F1203	Life Threatening Illness or Injury	A24	No Apparent Device Problem			
					CT257	Implantable cardiac pacemakers	E011903	Syncope/Fainting	A24	No Apparent Device Problem			
							E0602	Cardiac Arrest	A24	No Apparent Device Problem			
							E0604	Cardiac Perforation	A24	No Apparent Device Problem			
							E0605	Cardiac Tamponade	A24	No Apparent Device Problem			
							F08	Hospitalization or Prolonged Hospitalization	A24	No Apparent Device Problem			
							F1203	Life Threatening Illness or Injury	A24	No Apparent Device Problem			
					CT2727	Active implant devices/ancillaries	E011903	Syncope/Fainting	A24	No Apparent Device Problem			
							E0602	Cardiac Arrest	A24	No Apparent Device Problem			
							E0604	Cardiac Perforation	A24	No Apparent Device Problem			
							E0605	Cardiac Tamponade	A24	No Apparent Device Problem			
							F08	Hospitalization or Prolonged Hospitalization	A24	No Apparent Device Problem			
							F1203	Life Threatening Illness or Injury	A24	No Apparent Device Problem			
					CT2750	Implantable cardiac electrical stimulators and associated devices	E011903	Syncope/Fainting	A24	No Apparent Device Problem			
							E0602	Cardiac Arrest	A24	No Apparent Device Problem			
							E0604	Cardiac Perforation	A24	No Apparent Device Problem			
							E0605	Cardiac Tamponade	A24	No Apparent Device Problem			
							F08	Hospitalization or Prolonged Hospitalization	A24	No Apparent Device Problem			
							F1203	Life Threatening Illness or Injury	A24	No Apparent Device Problem			
					CT2818	Catheterization/Cannulation/Fluid path devices	E011903	Syncope/Fainting	A24	No Apparent Device Problem			
							E0602	Cardiac Arrest	A24	No Apparent Device Problem			
							E0604	Cardiac Perforation	A24	No Apparent Device Problem			
							E0605	Cardiac Tamponade	A24	No Apparent Device Problem			
							F08	Hospitalization or Prolonged Hospitalization	A24	No Apparent Device Problem			
							F1203	Life Threatening Illness or Injury	A24	No Apparent Device Problem			
					CT2866	Implantable cardiac electrical stimulators	E011903	Syncope/Fainting	A24	No Apparent Device Problem			
							E0602	Cardiac Arrest	A24	No Apparent Device Problem			
							E0604	Cardiac Perforation	A24	No Apparent Device Problem			

							E0605	Cardiac Tamponade	A24	No Apparent Device Problem		
							F08	Hospitalization or Prolonged Hospitalization	A24	No Apparent Device Problem		
							F1203	Life Threatening Illness or Injury	A24	No Apparent Device Problem		
					CT2985	Active	E011903	Syncope/Fainting	A24	No Apparent Device Problem		
							E0602	Cardiac Arrest	A24	No Apparent Device Problem		
							E0604	Cardiac Perforation	A24	No Apparent Device Problem		
							E0605	Cardiac Tamponade	A24	No Apparent Device Problem		
							F08	Hospitalization or Prolonged Hospitalization	A24	No Apparent Device Problem		
							F1203	Life Threatening Illness or Injury	A24	No Apparent Device Problem		
					CT3001	Implantable metal	E011903	Syncope/Fainting	A24	No Apparent Device Problem		
							E0602	Cardiac Arrest	A24	No Apparent Device Problem		
							E0604	Cardiac Perforation	A24	No Apparent Device Problem		
							E0605	Cardiac Tamponade	A24	No Apparent Device Problem		
							F08	Hospitalization or Prolonged Hospitalization	A24	No Apparent Device Problem		
							F1203	Life Threatening Illness or Injury	A24	No Apparent Device Problem		
					CT3002	Implantable synthetic polymer	E011903	Syncope/Fainting	A24	No Apparent Device Problem		
							E0602	Cardiac Arrest	A24	No Apparent Device Problem		
							E0604	Cardiac Perforation	A24	No Apparent Device Problem		
							E0605	Cardiac Tamponade	A24	No Apparent Device Problem		
							F08	Hospitalization or Prolonged Hospitalization	A24	No Apparent Device Problem		
							F1203	Life Threatening Illness or Injury	A24	No Apparent Device Problem		
					CT3006	Implantable inorganic material	E011903	Syncope/Fainting	A24	No Apparent Device Problem		
							E0602	Cardiac Arrest	A24	No Apparent Device Problem		
							E0604	Cardiac Perforation	A24	No Apparent Device Problem		
							E0605	Cardiac Tamponade	A24	No Apparent Device Problem		
							F08	Hospitalization or Prolonged Hospitalization	A24	No Apparent Device Problem		
							F1203	Life Threatening Illness or Injury	A24	No Apparent Device Problem		
					CT314	Electrically powered	E011903	Syncope/Fainting	A24	No Apparent Device Problem		
							E0602	Cardiac Arrest	A24	No Apparent Device Problem		
							E0604	Cardiac Perforation	A24	No Apparent Device Problem		
							E0605	Cardiac Tamponade	A24	No Apparent Device Problem		
							F08	Hospitalization or Prolonged Hospitalization	A24	No Apparent Device Problem		
							F1203	Life Threatening Illness or Injury	A24	No Apparent Device Problem		
					CT321	Long-term surgical invasive	E011903	Syncope/Fainting	A24	No Apparent Device Problem		
							E0602	Cardiac Arrest	A24	No Apparent Device Problem		
							E0604	Cardiac Perforation	A24	No Apparent Device Problem		
							E0605	Cardiac Tamponade	A24	No Apparent Device Problem		
							F08	Hospitalization or Prolonged Hospitalization	A24	No Apparent Device Problem		
							F1203	Life Threatening Illness or Injury	A24	No Apparent Device Problem		
					CT981	Single use	E011903	Syncope/Fainting	A24	No Apparent Device Problem		
							E0602	Cardiac Arrest	A24	No Apparent Device Problem		
							E0604	Cardiac Perforation	A24	No Apparent Device Problem		
							E0605	Cardiac Tamponade	A24	No Apparent Device Problem		
							F08	Hospitalization or Prolonged Hospitalization	A24	No Apparent Device Problem		
							F1203	Life Threatening Illness or Injury	A24	No Apparent Device Problem		
					CT983	Surgical invasive	E011903	Syncope/Fainting	A24	No Apparent Device Problem		
							E0602	Cardiac Arrest	A24	No Apparent Device Problem		
							E0604	Cardiac Perforation	A24	No Apparent Device Problem		
							E0605	Cardiac Tamponade	A24	No Apparent Device Problem		
							F08	Hospitalization or Prolonged Hospitalization	A24	No Apparent Device Problem		
							F1203	Life Threatening Illness or Injury	A24	No Apparent Device Problem		
9925	2020	All other reportable incidents	35852	Single-chamber implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	11/09/2020	11/09/2020
							F1901	Additional Surgery	A070504	Premature Discharge of Battery	11/09/2020	11/09/2020
9926	2020	All other reportable incidents	47267	Single-chamber implantable pacemaker, rate-responsive	CT141	Electrodes/Leads and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	24/08/2020	24/08/2020
							F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A070504	Premature Discharge of Battery	24/08/2020	24/08/2020
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	24/08/2020	24/08/2020
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery	24/08/2020	24/08/2020
					CT145	Cardiovascular devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	24/08/2020	24/08/2020
							F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A070504	Premature Discharge of Battery	24/08/2020	24/08/2020
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	24/08/2020	24/08/2020
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery	24/08/2020	24/08/2020
					CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	24/08/2020	24/08/2020
							F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A070504	Premature Discharge of Battery	24/08/2020	24/08/2020
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	24/08/2020	24/08/2020
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery	24/08/2020	24/08/2020
					CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	24/08/2020	24/08/2020
							F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A070504	Premature Discharge of Battery	24/08/2020	24/08/2020
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	24/08/2020	24/08/2020
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery	24/08/2020	24/08/2020
					CT233	Surgical	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	24/08/2020	24/08/2020
							F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A070504	Premature Discharge of Battery	24/08/2020	24/08/2020
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	24/08/2020	24/08/2020
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery	24/08/2020	24/08/2020
					CT2406	Implantable	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	24/08/2020	24/08/2020
							F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A070504	Premature Discharge of Battery	24/08/2020	24/08/2020
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	24/08/2020	24/08/2020
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery	24/08/2020	24/08/2020
					CT2407	Active implantable	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	24/08/2020	24/08/2020
							F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A070504	Premature Discharge of Battery	24/08/2020	24/08/2020
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	24/08/2020	24/08/2020
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery	24/08/2020	24/08/2020
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	24/08/2020	24/08/2020
							F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A070504	Premature Discharge of Battery	24/08/2020	24/08/2020
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	24/08/2020	24/08/2020
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery	24/08/2020	24/08/2020
					CT2727	Active implant devices/ancillaries	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	24/08/2020	24/08/2020
							F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A070504	Premature Discharge of Battery	24/08/2020	24/08/2020
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	24/08/2020	24/08/2020

							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery	24/08/2020	24/08/2020
					CT2750	Implantable cardiac electrical stimulators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	24/08/2020	24/08/2020
							F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A070504	Premature Discharge of Battery	24/08/2020	24/08/2020
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	24/08/2020	24/08/2020
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery	24/08/2020	24/08/2020
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	24/08/2020	24/08/2020
							F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A070504	Premature Discharge of Battery	24/08/2020	24/08/2020
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	24/08/2020	24/08/2020
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery	24/08/2020	24/08/2020
					CT2985	Active	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	24/08/2020	24/08/2020
							F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A070504	Premature Discharge of Battery	24/08/2020	24/08/2020
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	24/08/2020	24/08/2020
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery	24/08/2020	24/08/2020
					CT3001	Implantable metal	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	24/08/2020	24/08/2020
							F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A070504	Premature Discharge of Battery	24/08/2020	24/08/2020
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	24/08/2020	24/08/2020
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery	24/08/2020	24/08/2020
					CT3002	Implantable synthetic polymer	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	24/08/2020	24/08/2020
							F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A070504	Premature Discharge of Battery	24/08/2020	24/08/2020
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	24/08/2020	24/08/2020
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery	24/08/2020	24/08/2020
					CT3006	Implantable inorganic material	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	24/08/2020	24/08/2020
							F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A070504	Premature Discharge of Battery	24/08/2020	24/08/2020
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	24/08/2020	24/08/2020
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery	24/08/2020	24/08/2020
					CT314	Electrically powered	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	24/08/2020	24/08/2020
							F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A070504	Premature Discharge of Battery	24/08/2020	24/08/2020
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	24/08/2020	24/08/2020
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery	24/08/2020	24/08/2020
					CT321	Long-term surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	24/08/2020	24/08/2020
							F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A070504	Premature Discharge of Battery	24/08/2020	24/08/2020
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	24/08/2020	24/08/2020
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery	24/08/2020	24/08/2020
					CT449	Generators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	24/08/2020	24/08/2020
							F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A070504	Premature Discharge of Battery	24/08/2020	24/08/2020
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	24/08/2020	24/08/2020
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery	24/08/2020	24/08/2020
					CT981	Single use	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	24/08/2020	24/08/2020
							F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A070504	Premature Discharge of Battery	24/08/2020	24/08/2020
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	24/08/2020	24/08/2020
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery	24/08/2020	24/08/2020
					CT983	Surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	24/08/2020	24/08/2020
							F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A070504	Premature Discharge of Battery	24/08/2020	24/08/2020
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	24/08/2020	24/08/2020
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery	24/08/2020	24/08/2020
9927	2020	All other reportable incidents	17882	Physiologic-monitoring defibrillation system	CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070803	Failure to Power Up		
							F27	Problem Identified During Non-Clinical Procedure	A070803	Failure to Power Up		
9928	2020	All other reportable incidents	35852	Single-chamber implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E2104	Electric Shock	A0709	Device Sensing Problem		
									A071302	Inappropriate/Inadequate Shock/Stimulation		
									A1301	Failure to Read Input Signal		
							E2403	No Clinical Signs, Symptoms or Conditions	A0709	Device Sensing Problem		
									A071302	Inappropriate/Inadequate Shock/Stimulation		
									A1301	Failure to Read Input Signal		
							F26	No Health Consequences or Impact	A0709	Device Sensing Problem		
									A071302	Inappropriate/Inadequate Shock/Stimulation		
									A1301	Failure to Read Input Signal		
9929	2019	All other reportable incidents	37265	Dual-chamber implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A071302	Inappropriate/Inadequate Shock/Stimulation		
9930	2019	All other reportable incidents	37265	Dual-chamber implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A070909	Oversensing		
9931	2020	All other reportable incidents	37265	Dual-chamber implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A0508	Noise, Audible	19/11/2020	19/11/2020
									A070504	Premature Discharge of Battery	19/11/2020	19/11/2020
							F19	Surgical Intervention	A0508	Noise, Audible	19/11/2020	19/11/2020
									A070504	Premature Discharge of Battery	19/11/2020	19/11/2020
9932	2020	All other reportable incidents	47267	Single-chamber implantable pacemaker, rate-responsive	CT141	Electrodes/Leads and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	03/09/2020	03/09/2020
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	03/09/2020	03/09/2020
							F14	Prolonged Episode of Care	A070504	Premature Discharge of Battery	03/09/2020	03/09/2020
							F19	Surgical Intervention	A070504	Premature Discharge of Battery	03/09/2020	03/09/2020
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	03/09/2020	03/09/2020
					CT145	Cardiovascular devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	03/09/2020	03/09/2020
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	03/09/2020	03/09/2020
							F14	Prolonged Episode of Care	A070504	Premature Discharge of Battery	03/09/2020	03/09/2020
							F19	Surgical Intervention	A070504	Premature Discharge of Battery	03/09/2020	03/09/2020
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	03/09/2020	03/09/2020
					CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	03/09/2020	03/09/2020
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	03/09/2020	03/09/2020
							F14	Prolonged Episode of Care	A070504	Premature Discharge of Battery	03/09/2020	03/09/2020
							F19	Surgical Intervention	A070504	Premature Discharge of Battery	03/09/2020	03/09/2020
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	03/09/2020	03/09/2020
					CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	03/09/2020	03/09/2020
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	03/09/2020	03/09/2020
							F14	Prolonged Episode of Care	A070504	Premature Discharge of Battery	03/09/2020	03/09/2020
							F19	Surgical Intervention	A070504	Premature Discharge of Battery	03/09/2020	03/09/2020
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	03/09/2020	03/09/2020
					CT233	Surgical	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	03/09/2020	03/09/2020
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	03/09/2020	03/09/2020
							F14	Prolonged Episode of Care	A070504	Premature Discharge of Battery	03/09/2020	03/09/2020
							F19	Surgical Intervention	A070504	Premature Discharge of Battery	03/09/2020	03/09/2020
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	03/09/2020	03/09/2020
					CT2406	Implantable	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	03/09/2020	03/09/2020

							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	03/09/2020	03/09/2020
							F14	Prolonged Episode of Care	A070504	Premature Discharge of Battery	03/09/2020	03/09/2020
							F19	Surgical Intervention	A070504	Premature Discharge of Battery	03/09/2020	03/09/2020
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	03/09/2020	03/09/2020
					CT2407	Active implantable	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	03/09/2020	03/09/2020
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	03/09/2020	03/09/2020
							F14	Prolonged Episode of Care	A070504	Premature Discharge of Battery	03/09/2020	03/09/2020
							F19	Surgical Intervention	A070504	Premature Discharge of Battery	03/09/2020	03/09/2020
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	03/09/2020	03/09/2020
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	03/09/2020	03/09/2020
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	03/09/2020	03/09/2020
							F14	Prolonged Episode of Care	A070504	Premature Discharge of Battery	03/09/2020	03/09/2020
							F19	Surgical Intervention	A070504	Premature Discharge of Battery	03/09/2020	03/09/2020
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	03/09/2020	03/09/2020
					CT2727	Active implant devices/ancillaries	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	03/09/2020	03/09/2020
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	03/09/2020	03/09/2020
							F14	Prolonged Episode of Care	A070504	Premature Discharge of Battery	03/09/2020	03/09/2020
							F19	Surgical Intervention	A070504	Premature Discharge of Battery	03/09/2020	03/09/2020
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	03/09/2020	03/09/2020
					CT2750	Implantable cardiac electrical stimulators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	03/09/2020	03/09/2020
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	03/09/2020	03/09/2020
							F14	Prolonged Episode of Care	A070504	Premature Discharge of Battery	03/09/2020	03/09/2020
							F19	Surgical Intervention	A070504	Premature Discharge of Battery	03/09/2020	03/09/2020
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	03/09/2020	03/09/2020
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	03/09/2020	03/09/2020
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	03/09/2020	03/09/2020
							F14	Prolonged Episode of Care	A070504	Premature Discharge of Battery	03/09/2020	03/09/2020
							F19	Surgical Intervention	A070504	Premature Discharge of Battery	03/09/2020	03/09/2020
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	03/09/2020	03/09/2020
					CT2985	Active	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	03/09/2020	03/09/2020
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	03/09/2020	03/09/2020
							F14	Prolonged Episode of Care	A070504	Premature Discharge of Battery	03/09/2020	03/09/2020
							F19	Surgical Intervention	A070504	Premature Discharge of Battery	03/09/2020	03/09/2020
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	03/09/2020	03/09/2020
					CT3001	Implantable metal	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	03/09/2020	03/09/2020
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	03/09/2020	03/09/2020
							F14	Prolonged Episode of Care	A070504	Premature Discharge of Battery	03/09/2020	03/09/2020
							F19	Surgical Intervention	A070504	Premature Discharge of Battery	03/09/2020	03/09/2020
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	03/09/2020	03/09/2020
					CT3002	Implantable synthetic polymer	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	03/09/2020	03/09/2020
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	03/09/2020	03/09/2020
							F14	Prolonged Episode of Care	A070504	Premature Discharge of Battery	03/09/2020	03/09/2020
							F19	Surgical Intervention	A070504	Premature Discharge of Battery	03/09/2020	03/09/2020
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	03/09/2020	03/09/2020
					CT3006	Implantable inorganic material	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	03/09/2020	03/09/2020
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	03/09/2020	03/09/2020
							F14	Prolonged Episode of Care	A070504	Premature Discharge of Battery	03/09/2020	03/09/2020
							F19	Surgical Intervention	A070504	Premature Discharge of Battery	03/09/2020	03/09/2020
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	03/09/2020	03/09/2020
					CT314	Electrically powered	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	03/09/2020	03/09/2020
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	03/09/2020	03/09/2020
							F14	Prolonged Episode of Care	A070504	Premature Discharge of Battery	03/09/2020	03/09/2020
							F19	Surgical Intervention	A070504	Premature Discharge of Battery	03/09/2020	03/09/2020
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	03/09/2020	03/09/2020
					CT321	Long-term surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	03/09/2020	03/09/2020
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	03/09/2020	03/09/2020
							F14	Prolonged Episode of Care	A070504	Premature Discharge of Battery	03/09/2020	03/09/2020
							F19	Surgical Intervention	A070504	Premature Discharge of Battery	03/09/2020	03/09/2020
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	03/09/2020	03/09/2020
					CT449	Generators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	03/09/2020	03/09/2020
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	03/09/2020	03/09/2020
							F14	Prolonged Episode of Care	A070504	Premature Discharge of Battery	03/09/2020	03/09/2020
							F19	Surgical Intervention	A070504	Premature Discharge of Battery	03/09/2020	03/09/2020
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	03/09/2020	03/09/2020
					CT981	Single use	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	03/09/2020	03/09/2020
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	03/09/2020	03/09/2020
							F14	Prolonged Episode of Care	A070504	Premature Discharge of Battery	03/09/2020	03/09/2020
							F19	Surgical Intervention	A070504	Premature Discharge of Battery	03/09/2020	03/09/2020
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	03/09/2020	03/09/2020
					CT983	Surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	03/09/2020	03/09/2020
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	03/09/2020	03/09/2020
							F14	Prolonged Episode of Care	A070504	Premature Discharge of Battery	03/09/2020	03/09/2020
							F19	Surgical Intervention	A070504	Premature Discharge of Battery	03/09/2020	03/09/2020
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	03/09/2020	03/09/2020
9933	2020	All other reportable incidents	36241	Pacemaker/defibrillator lead	CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070101	Failure to Capture		
									A072201	High Impedance		
							F24	Insufficient Health Impact Information	A070101	Failure to Capture		
									A072201	High Impedance		
9934	2020	All other reportable incidents	47270	Cardiac resynchronization therapy implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A070101	Failure to Capture		
									A072201	High Impedance		
							F24	Insufficient Health Impact Information	A070101	Failure to Capture		
									A072201	High Impedance		
9935	2020	All other reportable incidents	37265	Dual-chamber implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	20/06/2022	20/06/2022
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	20/06/2022	20/06/2022
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	20/06/2022	20/06/2022
9936	2020	All other reportable incidents	37265	Dual-chamber implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A070909	Oversensing		
									A0712	Pacing Problem		
							F24	Insufficient Health Impact Information	A070909	Oversensing		

									A0712	Pacing Problem		
9937	2019	All other reportable incidents	47263	Cardiac resynchronization therapy implantable pacemaker	CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	30/12/2019	30/12/2019
					CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	30/12/2019	30/12/2019
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	30/12/2019	30/12/2019
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	30/12/2019	30/12/2019
9938	2020	All other reportable incidents	47265	Dual-chamber implantable pacemaker, rate-responsive	CT141	Electrodes/Leads and associated devices	E060109	Tachycardia	A0712	Pacing Problem		
							F11	Non-Serious Injury/ Illness/ Impairment	A0712	Pacing Problem		
					CT145	Cardiovascular devices	E060109	Tachycardia	A0712	Pacing Problem		
							F11	Non-Serious Injury/ Illness/ Impairment	A0712	Pacing Problem		
					CT1684	Cardiac pacemakers and associated devices	E060109	Tachycardia	A0712	Pacing Problem		
							F11	Non-Serious Injury/ Illness/ Impairment	A0712	Pacing Problem		
					CT1685	Cardiac pacemakers	E060109	Tachycardia	A0712	Pacing Problem		
							F11	Non-Serious Injury/ Illness/ Impairment	A0712	Pacing Problem		
					CT233	Surgical	E060109	Tachycardia	A0712	Pacing Problem		
							F11	Non-Serious Injury/ Illness/ Impairment	A0712	Pacing Problem		
					CT2406	Implantable	E060109	Tachycardia	A0712	Pacing Problem		
							F11	Non-Serious Injury/ Illness/ Impairment	A0712	Pacing Problem		
					CT2407	Active implantable	E060109	Tachycardia	A0712	Pacing Problem		
							F11	Non-Serious Injury/ Illness/ Impairment	A0712	Pacing Problem		
					CT257	Implantable cardiac pacemakers	E060109	Tachycardia	A0712	Pacing Problem		
							F11	Non-Serious Injury/ Illness/ Impairment	A0712	Pacing Problem		
					CT2727	Active implant devices/ancillaries	E060109	Tachycardia	A0712	Pacing Problem		
							F11	Non-Serious Injury/ Illness/ Impairment	A0712	Pacing Problem		
					CT2750	Implantable cardiac electrical stimulators and associated devices	E060109	Tachycardia	A0712	Pacing Problem		
							F11	Non-Serious Injury/ Illness/ Impairment	A0712	Pacing Problem		
					CT2866	Implantable cardiac electrical stimulators	E060109	Tachycardia	A0712	Pacing Problem		
							F11	Non-Serious Injury/ Illness/ Impairment	A0712	Pacing Problem		
					CT2985	Active	E060109	Tachycardia	A0712	Pacing Problem		
							F11	Non-Serious Injury/ Illness/ Impairment	A0712	Pacing Problem		
					CT3001	Implantable metal	E060109	Tachycardia	A0712	Pacing Problem		
							F11	Non-Serious Injury/ Illness/ Impairment	A0712	Pacing Problem		
					CT3002	Implantable synthetic polymer	E060109	Tachycardia	A0712	Pacing Problem		
							F11	Non-Serious Injury/ Illness/ Impairment	A0712	Pacing Problem		
					CT3006	Implantable inorganic material	E060109	Tachycardia	A0712	Pacing Problem		
							F11	Non-Serious Injury/ Illness/ Impairment	A0712	Pacing Problem		
					CT314	Electrically powered	E060109	Tachycardia	A0712	Pacing Problem		
							F11	Non-Serious Injury/ Illness/ Impairment	A0712	Pacing Problem		
					CT321	Long-term surgical invasive	E060109	Tachycardia	A0712	Pacing Problem		
							F11	Non-Serious Injury/ Illness/ Impairment	A0712	Pacing Problem		
					CT449	Generators and associated devices	E060109	Tachycardia	A0712	Pacing Problem		
							F11	Non-Serious Injury/ Illness/ Impairment	A0712	Pacing Problem		
					CT981	Single use	E060109	Tachycardia	A0712	Pacing Problem		
							F11	Non-Serious Injury/ Illness/ Impairment	A0712	Pacing Problem		
					CT983	Surgical invasive	E060109	Tachycardia	A0712	Pacing Problem		
							F11	Non-Serious Injury/ Illness/ Impairment	A0712	Pacing Problem		
9939	2020	All other reportable incidents	60910	Coronary venous pacing lead	CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070101	Failure to Capture		
									A0722	Impedance Problem		
									A072201	High Impedance		
							F26	No Health Consequences or Impact	A070101	Failure to Capture		
									A0722	Impedance Problem		
									A072201	High Impedance		
9940	2020	All other reportable incidents	37265	Dual-chamber implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E0601	Arrhythmia	A071202	Inaccurate Synchronization		
							E0602	Cardiac Arrest	A071202	Inaccurate Synchronization		
							E2403	No Clinical Signs, Symptoms or Conditions	A071202	Inaccurate Synchronization		
							F12	Serious Injury/ Illness/ Impairment	A071202	Inaccurate Synchronization		
9941	2020	All other reportable incidents	47265	Dual-chamber implantable pacemaker, rate-responsive	CT141	Electrodes/Leads and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A180104	Device Contamination with Chemical or Other Material	12/04/2019	12/04/2019
							F26	No Health Consequences or Impact	A180104	Device Contamination with Chemical or Other Material	12/04/2019	12/04/2019
					CT145	Cardiovascular devices	E2403	No Clinical Signs, Symptoms or Conditions	A180104	Device Contamination with Chemical or Other Material	12/04/2019	12/04/2019
							F26	No Health Consequences or Impact	A180104	Device Contamination with Chemical or Other Material	12/04/2019	12/04/2019
					CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A180104	Device Contamination with Chemical or Other Material	12/04/2019	12/04/2019
							F26	No Health Consequences or Impact	A180104	Device Contamination with Chemical or Other Material	12/04/2019	12/04/2019
					CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A180104	Device Contamination with Chemical or Other Material	12/04/2019	12/04/2019
							F26	No Health Consequences or Impact	A180104	Device Contamination with Chemical or Other Material	12/04/2019	12/04/2019
					CT233	Surgical	E2403	No Clinical Signs, Symptoms or Conditions	A180104	Device Contamination with Chemical or Other Material	12/04/2019	12/04/2019
							F26	No Health Consequences or Impact	A180104	Device Contamination with Chemical or Other Material	12/04/2019	12/04/2019
					CT2406	Implantable	E2403	No Clinical Signs, Symptoms or Conditions	A180104	Device Contamination with Chemical or Other Material	12/04/2019	12/04/2019
							F26	No Health Consequences or Impact	A180104	Device Contamination with Chemical or Other Material	12/04/2019	12/04/2019
					CT2407	Active implantable	E2403	No Clinical Signs, Symptoms or Conditions	A180104	Device Contamination with Chemical or Other Material	12/04/2019	12/04/2019
							F26	No Health Consequences or Impact	A180104	Device Contamination with Chemical or Other Material	12/04/2019	12/04/2019
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A180104	Device Contamination with Chemical or Other Material	12/04/2019	12/04/2019
							F26	No Health Consequences or Impact	A180104	Device Contamination with Chemical or Other Material	12/04/2019	12/04/2019
					CT2727	Active implant devices/ancillaries	E2403	No Clinical Signs, Symptoms or Conditions	A180104	Device Contamination with Chemical or Other Material	12/04/2019	12/04/2019
							F26	No Health Consequences or Impact	A180104	Device Contamination with Chemical or Other Material	12/04/2019	12/04/2019

					CT2750	Implantable cardiac electrical stimulators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A180104	Device Contamination with Chemical or Other Material	12/04/2019	12/04/2019
							F26	No Health Consequences or Impact	A180104	Device Contamination with Chemical or Other Material	12/04/2019	12/04/2019
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A180104	Device Contamination with Chemical or Other Material	12/04/2019	12/04/2019
							F26	No Health Consequences or Impact	A180104	Device Contamination with Chemical or Other Material	12/04/2019	12/04/2019
					CT2985	Active	E2403	No Clinical Signs, Symptoms or Conditions	A180104	Device Contamination with Chemical or Other Material	12/04/2019	12/04/2019
							F26	No Health Consequences or Impact	A180104	Device Contamination with Chemical or Other Material	12/04/2019	12/04/2019
					CT3001	Implantable metal	E2403	No Clinical Signs, Symptoms or Conditions	A180104	Device Contamination with Chemical or Other Material	12/04/2019	12/04/2019
							F26	No Health Consequences or Impact	A180104	Device Contamination with Chemical or Other Material	12/04/2019	12/04/2019
					CT3002	Implantable synthetic polymer	E2403	No Clinical Signs, Symptoms or Conditions	A180104	Device Contamination with Chemical or Other Material	12/04/2019	12/04/2019
							F26	No Health Consequences or Impact	A180104	Device Contamination with Chemical or Other Material	12/04/2019	12/04/2019
					CT3006	Implantable inorganic material	E2403	No Clinical Signs, Symptoms or Conditions	A180104	Device Contamination with Chemical or Other Material	12/04/2019	12/04/2019
							F26	No Health Consequences or Impact	A180104	Device Contamination with Chemical or Other Material	12/04/2019	12/04/2019
					CT314	Electrically powered	E2403	No Clinical Signs, Symptoms or Conditions	A180104	Device Contamination with Chemical or Other Material	12/04/2019	12/04/2019
							F26	No Health Consequences or Impact	A180104	Device Contamination with Chemical or Other Material	12/04/2019	12/04/2019
					CT321	Long-term surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A180104	Device Contamination with Chemical or Other Material	12/04/2019	12/04/2019
							F26	No Health Consequences or Impact	A180104	Device Contamination with Chemical or Other Material	12/04/2019	12/04/2019
					CT449	Generators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A180104	Device Contamination with Chemical or Other Material	12/04/2019	12/04/2019
							F26	No Health Consequences or Impact	A180104	Device Contamination with Chemical or Other Material	12/04/2019	12/04/2019
					CT981	Single use	E2403	No Clinical Signs, Symptoms or Conditions	A180104	Device Contamination with Chemical or Other Material	12/04/2019	12/04/2019
							F26	No Health Consequences or Impact	A180104	Device Contamination with Chemical or Other Material	12/04/2019	12/04/2019
					CT983	Surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A180104	Device Contamination with Chemical or Other Material	12/04/2019	12/04/2019
							F26	No Health Consequences or Impact	A180104	Device Contamination with Chemical or Other Material	12/04/2019	12/04/2019
9942	2020	All other reportable incidents	36241	Pacemaker/defibrillator lead	CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A0203	Defective Device		
									A070909	Oversensing		
									A090801	Signal Artifact/Noise		
							F24	Insufficient Health Impact Information	A0203	Defective Device		
									A070909	Oversensing		
									A090801	Signal Artifact/Noise		
9943	2020	All other reportable incidents	36241	Pacemaker/defibrillator lead	CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070909	Oversensing		
									A090801	Signal Artifact/Noise		
							F26	No Health Consequences or Impact	A070909	Oversensing		
									A090801	Signal Artifact/Noise		
9944	2020	All other reportable incidents	47265	Dual-chamber implantable pacemaker, rate-responsive	CT141	Electrodes/Leads and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070909	Oversensing		
							F26	No Health Consequences or Impact	A070909	Oversensing		
					CT145	Cardiovascular devices	E2403	No Clinical Signs, Symptoms or Conditions	A070909	Oversensing		
							F26	No Health Consequences or Impact	A070909	Oversensing		
					CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070909	Oversensing		
							F26	No Health Consequences or Impact	A070909	Oversensing		
					CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A070909	Oversensing		
							F26	No Health Consequences or Impact	A070909	Oversensing		
					CT233	Surgical	E2403	No Clinical Signs, Symptoms or Conditions	A070909	Oversensing		
							F26	No Health Consequences or Impact	A070909	Oversensing		
					CT2406	Implantable	E2403	No Clinical Signs, Symptoms or Conditions	A070909	Oversensing		
							F26	No Health Consequences or Impact	A070909	Oversensing		
					CT2407	Active implantable	E2403	No Clinical Signs, Symptoms or Conditions	A070909	Oversensing		
							F26	No Health Consequences or Impact	A070909	Oversensing		
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A070909	Oversensing		
							F26	No Health Consequences or Impact	A070909	Oversensing		
					CT2727	Active implant devices/ancillaries	E2403	No Clinical Signs, Symptoms or Conditions	A070909	Oversensing		
							F26	No Health Consequences or Impact	A070909	Oversensing		
					CT2750	Implantable cardiac electrical stimulators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070909	Oversensing		
							F26	No Health Consequences or Impact	A070909	Oversensing		
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A070909	Oversensing		
							F26	No Health Consequences or Impact	A070909	Oversensing		
					CT2985	Active	E2403	No Clinical Signs, Symptoms or Conditions	A070909	Oversensing		
							F26	No Health Consequences or Impact	A070909	Oversensing		
					CT3001	Implantable metal	E2403	No Clinical Signs, Symptoms or Conditions	A070909	Oversensing		
							F26	No Health Consequences or Impact	A070909	Oversensing		
					CT3002	Implantable synthetic polymer	E2403	No Clinical Signs, Symptoms or Conditions	A070909	Oversensing		
							F26	No Health Consequences or Impact	A070909	Oversensing		
					CT3006	Implantable inorganic material	E2403	No Clinical Signs, Symptoms or Conditions	A070909	Oversensing		
							F26	No Health Consequences or Impact	A070909	Oversensing		
					CT314	Electrically powered	E2403	No Clinical Signs, Symptoms or Conditions	A070909	Oversensing		
							F26	No Health Consequences or Impact	A070909	Oversensing		
					CT321	Long-term surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A070909	Oversensing		
							F26	No Health Consequences or Impact	A070909	Oversensing		
					CT449	Generators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070909	Oversensing		
							F26	No Health Consequences or Impact	A070909	Oversensing		
					CT981	Single use	E2403	No Clinical Signs, Symptoms or Conditions	A070909	Oversensing		
							F26	No Health Consequences or Impact	A070909	Oversensing		
					CT983	Surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A070909	Oversensing		

							F26	No Health Consequences or Impact	A070909	Oversensing		
9945	2020	All other reportable incidents	36241	Pacemaker/defibrillator lead	CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070909	Oversensing		
							F26	No Health Consequences or Impact	A070909	Oversensing		
9946	2019	All other reportable incidents	35039	Epicardial pacing lead	CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A090801	Signal Artifact/Noise		
9947	2020	All other reportable incidents	47265	Dual-chamber implantable pacemaker, rate-responsive	CT141	Electrodes/Leads and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
					CT145	Cardiovascular devices	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
					CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
					CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
					CT233	Surgical	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
					CT2406	Implantable	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
					CT2407	Active implantable	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
					CT2727	Active implant devices/ancillaries	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
					CT2750	Implantable cardiac electrical stimulators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
					CT2985	Active	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
					CT3001	Implantable metal	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
					CT3002	Implantable synthetic polymer	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
					CT3006	Implantable inorganic material	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
					CT314	Electrically powered	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
					CT321	Long-term surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
					CT449	Generators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
					CT981	Single use	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
					CT983	Surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
9948	2020	All other reportable incidents	35852	Single-chamber implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
9949	2019	All other reportable incidents	47263	Cardiac resynchronization therapy implantable pacemaker	CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A071205	Pacemaker Found in Back-Up Mode		
					CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A071205	Pacemaker Found in Back-Up Mode		

					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A071205	Pacemaker Found in Back-Up Mode			
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A071205	Pacemaker Found in Back-Up Mode			
9950	2019	All other reportable incidents	47267	Single-chamber implantable pacemaker, rate-responsive	CT141	Electrodes/Leads and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery			
					CT145	Cardiovascular devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery			
					CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery			
					CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery			
					CT233	Surgical	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery			
					CT2406	Implantable	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery			
					CT2407	Active implantable	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery			
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery			
					CT2727	Active implant devices/ancillaries	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery			
					CT2750	Implantable cardiac electrical stimulators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery			
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery			
					CT2985	Active	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery			
					CT3001	Implantable metal	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery			
					CT3002	Implantable synthetic polymer	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery			
					CT3006	Implantable inorganic material	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery			
					CT314	Electrically powered	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery			
					CT321	Long-term surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery			
					CT449	Generators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery			
					CT981	Single use	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery			
					CT983	Surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery			
9951	2019	All other reportable incidents	36241	Pacemaker/defibrillator lead	CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A010402	Migration			
9952	2019	All other reportable incidents	35852	Single-chamber implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	F1001	Absence of Treatment	A070909	Oversensing			
9953	2019	All other reportable incidents	36241	Pacemaker/defibrillator lead	CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A0510	Retraction Problem	04/10/2019		
9954	2019	All other reportable incidents	47267	Single-chamber implantable pacemaker, rate-responsive	CT141	Electrodes/Leads and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A1606	Premature Indicator Activation	11/10/2019		
					CT145	Cardiovascular devices	E2403	No Clinical Signs, Symptoms or Conditions	A1606	Premature Indicator Activation	11/10/2019		
					CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A1606	Premature Indicator Activation	11/10/2019		
					CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A1606	Premature Indicator Activation	11/10/2019		
					CT233	Surgical	E2403	No Clinical Signs, Symptoms or Conditions	A1606	Premature Indicator Activation	11/10/2019		
					CT2406	Implantable	E2403	No Clinical Signs, Symptoms or Conditions	A1606	Premature Indicator Activation	11/10/2019		
					CT2407	Active implantable	E2403	No Clinical Signs, Symptoms or Conditions	A1606	Premature Indicator Activation	11/10/2019		
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A1606	Premature Indicator Activation	11/10/2019		
					CT2727	Active implant devices/ancillaries	E2403	No Clinical Signs, Symptoms or Conditions	A1606	Premature Indicator Activation	11/10/2019		
					CT2750	Implantable cardiac electrical stimulators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A1606	Premature Indicator Activation	11/10/2019		
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A1606	Premature Indicator Activation	11/10/2019		
					CT2985	Active	E2403	No Clinical Signs, Symptoms or Conditions	A1606	Premature Indicator Activation	11/10/2019		
					CT3001	Implantable metal	E2403	No Clinical Signs, Symptoms or Conditions	A1606	Premature Indicator Activation	11/10/2019		
					CT3002	Implantable synthetic polymer	E2403	No Clinical Signs, Symptoms or Conditions	A1606	Premature Indicator Activation	11/10/2019		
					CT3006	Implantable inorganic material	E2403	No Clinical Signs, Symptoms or Conditions	A1606	Premature Indicator Activation	11/10/2019		
					CT314	Electrically powered	E2403	No Clinical Signs, Symptoms or Conditions	A1606	Premature Indicator Activation	11/10/2019		
					CT321	Long-term surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A1606	Premature Indicator Activation	11/10/2019		
					CT449	Generators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A1606	Premature Indicator Activation	11/10/2019		
					CT981	Single use	E2403	No Clinical Signs, Symptoms or Conditions	A1606	Premature Indicator Activation	11/10/2019		
					CT983	Surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A1606	Premature Indicator Activation	11/10/2019		
9955	2019	All other reportable incidents	47263	Cardiac resynchronization therapy implantable pacemaker	CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery			
					CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery			
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery			
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery			
9956	2019	All other reportable incidents	36241	Pacemaker/defibrillator lead	CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A04	Material Integrity Problem	20/09/2019		
									A150206	Difficult to Insert	20/09/2019		
							F26	No Health Consequences or Impact	A04	Material Integrity Problem	20/09/2019		
									A150206	Difficult to Insert	20/09/2019		
9957	2020	All other reportable incidents	47267	Single-chamber implantable pacemaker, rate-responsive	CT141	Electrodes/Leads and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	21/09/2020	21/09/2020	
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	21/09/2020	21/09/2020	
							F12	Serious Injury/ Illness/ Impairment	A070504	Premature Discharge of Battery	21/09/2020	21/09/2020	
							F19	Surgical Intervention	A070504	Premature Discharge of Battery	21/09/2020	21/09/2020	
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	21/09/2020	21/09/2020	
					CT145	Cardiovascular devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	21/09/2020	21/09/2020	
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	21/09/2020	21/09/2020	
							F12	Serious Injury/ Illness/ Impairment	A070504	Premature Discharge of Battery	21/09/2020	21/09/2020	
							F19	Surgical Intervention	A070504	Premature Discharge of Battery	21/09/2020	21/09/2020	
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	21/09/2020	21/09/2020	
					CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	21/09/2020	21/09/2020	
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	21/09/2020	21/09/2020	
							F12	Serious Injury/ Illness/ Impairment	A070504	Premature Discharge of Battery	21/09/2020	21/09/2020	
							F19	Surgical Intervention	A070504	Premature Discharge of Battery	21/09/2020	21/09/2020	
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	21/09/2020	21/09/2020	
					CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	21/09/2020	21/09/2020	
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	21/09/2020	21/09/2020	
							F12	Serious Injury/ Illness/ Impairment	A070504	Premature Discharge of Battery	21/09/2020	21/09/2020	
							F19	Surgical Intervention	A070504	Premature Discharge of Battery	21/09/2020	21/09/2020	
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	21/09/2020	21/09/2020	
					CT233	Surgical	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	21/09/2020	21/09/2020	
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	21/09/2020	21/09/2020	
							F12	Serious Injury/ Illness/ Impairment	A070504	Premature Discharge of Battery	21/09/2020	21/09/2020	
							F19	Surgical Intervention	A070504	Premature Discharge of Battery	21/09/2020	21/09/2020	
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	21/09/2020	21/09/2020	
					CT2406	Implantable	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	21/09/2020	21/09/2020	
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	21/09/2020	21/09/2020	
							F12	Serious Injury/ Illness/ Impairment	A070504	Premature Discharge of Battery	21/09/2020	21/09/2020	
							F19	Surgical Intervention	A070504	Premature Discharge of Battery	21/09/2020	21/09/2020	
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	21/09/2020	21/09/2020	
					CT2407	Active implantable	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	21/09/2020	21/09/2020	
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	21/09/2020	21/09/2020	
							F12	Serious Injury/ Illness/ Impairment	A070504	Premature Discharge of Battery	21/09/2020	21/09/2020	

							F19	Surgical Intervention	A070504	Premature Discharge of Battery	21/09/2020	21/09/2020
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	21/09/2020	21/09/2020
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	21/09/2020	21/09/2020
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	21/09/2020	21/09/2020
							F12	Serious Injury/ Illness/ Impairment	A070504	Premature Discharge of Battery	21/09/2020	21/09/2020
							F19	Surgical Intervention	A070504	Premature Discharge of Battery	21/09/2020	21/09/2020
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	21/09/2020	21/09/2020
					CT2727	Active implant devices/ancillaries	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	21/09/2020	21/09/2020
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	21/09/2020	21/09/2020
							F12	Serious Injury/ Illness/ Impairment	A070504	Premature Discharge of Battery	21/09/2020	21/09/2020
							F19	Surgical Intervention	A070504	Premature Discharge of Battery	21/09/2020	21/09/2020
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	21/09/2020	21/09/2020
					CT2750	Implantable cardiac electrical stimulators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	21/09/2020	21/09/2020
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	21/09/2020	21/09/2020
							F12	Serious Injury/ Illness/ Impairment	A070504	Premature Discharge of Battery	21/09/2020	21/09/2020
							F19	Surgical Intervention	A070504	Premature Discharge of Battery	21/09/2020	21/09/2020
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	21/09/2020	21/09/2020
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	21/09/2020	21/09/2020
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	21/09/2020	21/09/2020
							F12	Serious Injury/ Illness/ Impairment	A070504	Premature Discharge of Battery	21/09/2020	21/09/2020
							F19	Surgical Intervention	A070504	Premature Discharge of Battery	21/09/2020	21/09/2020
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	21/09/2020	21/09/2020
					CT2985	Active	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	21/09/2020	21/09/2020
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	21/09/2020	21/09/2020
							F12	Serious Injury/ Illness/ Impairment	A070504	Premature Discharge of Battery	21/09/2020	21/09/2020
							F19	Surgical Intervention	A070504	Premature Discharge of Battery	21/09/2020	21/09/2020
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	21/09/2020	21/09/2020
					CT3001	Implantable metal	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	21/09/2020	21/09/2020
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	21/09/2020	21/09/2020
							F12	Serious Injury/ Illness/ Impairment	A070504	Premature Discharge of Battery	21/09/2020	21/09/2020
							F19	Surgical Intervention	A070504	Premature Discharge of Battery	21/09/2020	21/09/2020
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	21/09/2020	21/09/2020
					CT3002	Implantable synthetic polymer	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	21/09/2020	21/09/2020
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	21/09/2020	21/09/2020
							F12	Serious Injury/ Illness/ Impairment	A070504	Premature Discharge of Battery	21/09/2020	21/09/2020
							F19	Surgical Intervention	A070504	Premature Discharge of Battery	21/09/2020	21/09/2020
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	21/09/2020	21/09/2020
					CT3006	Implantable inorganic material	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	21/09/2020	21/09/2020
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	21/09/2020	21/09/2020
							F12	Serious Injury/ Illness/ Impairment	A070504	Premature Discharge of Battery	21/09/2020	21/09/2020
							F19	Surgical Intervention	A070504	Premature Discharge of Battery	21/09/2020	21/09/2020
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	21/09/2020	21/09/2020
					CT314	Electrically powered	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	21/09/2020	21/09/2020
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	21/09/2020	21/09/2020
							F12	Serious Injury/ Illness/ Impairment	A070504	Premature Discharge of Battery	21/09/2020	21/09/2020
							F19	Surgical Intervention	A070504	Premature Discharge of Battery	21/09/2020	21/09/2020
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	21/09/2020	21/09/2020
					CT321	Long-term surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	21/09/2020	21/09/2020
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	21/09/2020	21/09/2020
							F12	Serious Injury/ Illness/ Impairment	A070504	Premature Discharge of Battery	21/09/2020	21/09/2020
							F19	Surgical Intervention	A070504	Premature Discharge of Battery	21/09/2020	21/09/2020
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	21/09/2020	21/09/2020
					CT449	Generators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	21/09/2020	21/09/2020
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	21/09/2020	21/09/2020
							F12	Serious Injury/ Illness/ Impairment	A070504	Premature Discharge of Battery	21/09/2020	21/09/2020
							F19	Surgical Intervention	A070504	Premature Discharge of Battery	21/09/2020	21/09/2020
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	21/09/2020	21/09/2020
					CT981	Single use	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	21/09/2020	21/09/2020
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	21/09/2020	21/09/2020
							F12	Serious Injury/ Illness/ Impairment	A070504	Premature Discharge of Battery	21/09/2020	21/09/2020
							F19	Surgical Intervention	A070504	Premature Discharge of Battery	21/09/2020	21/09/2020
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	21/09/2020	21/09/2020
					CT983	Surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	21/09/2020	21/09/2020
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	21/09/2020	21/09/2020
							F12	Serious Injury/ Illness/ Impairment	A070504	Premature Discharge of Battery	21/09/2020	21/09/2020
							F19	Surgical Intervention	A070504	Premature Discharge of Battery	21/09/2020	21/09/2020
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	21/09/2020	21/09/2020
9958	2019	All other reportable incidents	35852	Single-chamber implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A0709	Device Sensing Problem		
9959	2020	All other reportable incidents	35852	Single-chamber implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	01/09/2020	01/09/2020
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	01/09/2020	01/09/2020
9960	2020	All other reportable incidents	47265	Dual-chamber implantable pacemaker, rate-responsive	CT141	Electrodes/Leads and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A090804	False Positive Result		
							F26	No Health Consequences or Impact	A090804	False Positive Result		
					CT145	Cardiovascular devices	E2403	No Clinical Signs, Symptoms or Conditions	A090804	False Positive Result		
							F26	No Health Consequences or Impact	A090804	False Positive Result		
					CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A090804	False Positive Result		
							F26	No Health Consequences or Impact	A090804	False Positive Result		
					CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A090804	False Positive Result		
							F26	No Health Consequences or Impact	A090804	False Positive Result		
					CT233	Surgical	E2403	No Clinical Signs, Symptoms or Conditions	A090804	False Positive Result		
							F26	No Health Consequences or Impact	A090804	False Positive Result		
					CT2406	Implantable	E2403	No Clinical Signs, Symptoms or Conditions	A090804	False Positive Result		
							F26	No Health Consequences or Impact	A090804	False Positive Result		
					CT2407	Active implantable	E2403	No Clinical Signs, Symptoms or Conditions	A090804	False Positive Result		
							F26	No Health Consequences or Impact	A090804	False Positive Result		
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A090804	False Positive Result		
							F26	No Health Consequences or Impact	A090804	False Positive Result		
					CT2727	Active implant devices/ancillaries	E2403	No Clinical Signs, Symptoms or Conditions	A090804	False Positive Result		
							F26	No Health Consequences or Impact	A090804	False Positive Result		

					CT2750	Implantable cardiac electrical stimulators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A090804	False Positive Result		
							F26	No Health Consequences or Impact	A090804	False Positive Result		
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A090804	False Positive Result		
							F26	No Health Consequences or Impact	A090804	False Positive Result		
					CT2985	Active	E2403	No Clinical Signs, Symptoms or Conditions	A090804	False Positive Result		
							F26	No Health Consequences or Impact	A090804	False Positive Result		
					CT3001	Implantable metal	E2403	No Clinical Signs, Symptoms or Conditions	A090804	False Positive Result		
							F26	No Health Consequences or Impact	A090804	False Positive Result		
					CT3002	Implantable synthetic polymer	E2403	No Clinical Signs, Symptoms or Conditions	A090804	False Positive Result		
							F26	No Health Consequences or Impact	A090804	False Positive Result		
					CT3006	Implantable inorganic material	E2403	No Clinical Signs, Symptoms or Conditions	A090804	False Positive Result		
							F26	No Health Consequences or Impact	A090804	False Positive Result		
					CT314	Electrically powered	E2403	No Clinical Signs, Symptoms or Conditions	A090804	False Positive Result		
							F26	No Health Consequences or Impact	A090804	False Positive Result		
					CT321	Long-term surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A090804	False Positive Result		
							F26	No Health Consequences or Impact	A090804	False Positive Result		
					CT449	Generators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A090804	False Positive Result		
							F26	No Health Consequences or Impact	A090804	False Positive Result		
					CT981	Single use	E2403	No Clinical Signs, Symptoms or Conditions	A090804	False Positive Result		
							F26	No Health Consequences or Impact	A090804	False Positive Result		
					CT983	Surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A090804	False Positive Result		
							F26	No Health Consequences or Impact	A090804	False Positive Result		
9961	2019	All other reportable incidents	35852	Single-chamber implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	02/10/2019	
9962	2020	All other reportable incidents	35852	Single-chamber implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E2104	Electric Shock	A071302	Inappropriate/Inadequate Shock/Stimulation		
							F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A071302	Inappropriate/Inadequate Shock/Stimulation		
							F26	No Health Consequences or Impact	A071302	Inappropriate/Inadequate Shock/Stimulation		
9963	2019	All other reportable incidents	37265	Dual-chamber implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	01/10/2019	
9964	2020	All other reportable incidents	47270	Cardiac resynchronization therapy implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E060110	Ventricular Fibrillation	A071202	Inaccurate Synchronization		
							E2403	No Clinical Signs, Symptoms or Conditions	A071202	Inaccurate Synchronization		
							F12	Serious Injury/ Illness/ Impairment	A071202	Inaccurate Synchronization		
9965	2019	All other reportable incidents	36241	Pacemaker/defibrillator lead	CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A05	Mechanical Problem	04/10/2019	04/10/2019
							F26	No Health Consequences or Impact	A05	Mechanical Problem	04/10/2019	04/10/2019
9966	2019	All other reportable incidents	35852	Single-chamber implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A072202	Low Impedance		
9967	2019	All other reportable incidents	35852	Single-chamber implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A072201	High Impedance		
9968	2019	All other reportable incidents	47263	Cardiac resynchronization therapy implantable pacemaker	CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
					CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
9969	2020	All other reportable incidents	47263	Cardiac resynchronization therapy implantable pacemaker	CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	09/09/2020	09/09/2020
							F1901	Additional Surgery	A070504	Premature Discharge of Battery	09/09/2020	09/09/2020
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	09/09/2020	09/09/2020
					CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	09/09/2020	09/09/2020
							F1901	Additional Surgery	A070504	Premature Discharge of Battery	09/09/2020	09/09/2020
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	09/09/2020	09/09/2020
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	09/09/2020	09/09/2020
							F1901	Additional Surgery	A070504	Premature Discharge of Battery	09/09/2020	09/09/2020
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	09/09/2020	09/09/2020
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	09/09/2020	09/09/2020
							F1901	Additional Surgery	A070504	Premature Discharge of Battery	09/09/2020	09/09/2020
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	09/09/2020	09/09/2020
9970	2019	Death	47270	Cardiac resynchronization therapy implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	F21	Unexpected Deterioration	A0712	Pacing Problem		
9971	2019	All other reportable incidents	37265	Dual-chamber implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	F1905	Device Revision or Replacement	A0705	Battery Problem	21/10/2019	
9972	2019	All other reportable incidents	37265	Dual-chamber implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	01/03/2020	31/03/2020
							F19	Surgical Intervention	A070504	Premature Discharge of Battery	01/03/2020	31/03/2020
9973	2019	All other reportable incidents	35852	Single-chamber implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A070909	Oversensing		
9974	2020	All other reportable incidents	47265	Dual-chamber implantable pacemaker, rate-responsive	CT141	Electrodes/Leads and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	11/09/2020	10/12/2020
							F19	Surgical Intervention	A070504	Premature Discharge of Battery	11/09/2020	10/12/2020
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	11/09/2020	10/12/2020
					CT145	Cardiovascular devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	11/09/2020	10/12/2020
							F19	Surgical Intervention	A070504	Premature Discharge of Battery	11/09/2020	10/12/2020
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	11/09/2020	10/12/2020
					CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	11/09/2020	10/12/2020
							F19	Surgical Intervention	A070504	Premature Discharge of Battery	11/09/2020	10/12/2020
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	11/09/2020	10/12/2020
					CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	11/09/2020	10/12/2020
							F19	Surgical Intervention	A070504	Premature Discharge of Battery	11/09/2020	10/12/2020
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	11/09/2020	10/12/2020
					CT233	Surgical	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	11/09/2020	10/12/2020
							F19	Surgical Intervention	A070504	Premature Discharge of Battery	11/09/2020	10/12/2020
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	11/09/2020	10/12/2020
					CT2406	Implantable	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	11/09/2020	10/12/2020
							F19	Surgical Intervention	A070504	Premature Discharge of Battery	11/09/2020	10/12/2020
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	11/09/2020	10/12/2020
					CT2407	Active implantable	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	11/09/2020	10/12/2020
							F19	Surgical Intervention	A070504	Premature Discharge of Battery	11/09/2020	10/12/2020
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	11/09/2020	10/12/2020
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	11/09/2020	10/12/2020
							F19	Surgical Intervention	A070504	Premature Discharge of Battery	11/09/2020	10/12/2020
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	11/09/2020	10/12/2020
					CT2727	Active implant devices/ancillaries	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	11/09/2020	10/12/2020
							F19	Surgical Intervention	A070504	Premature Discharge of Battery	11/09/2020	10/12/2020
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	11/09/2020	10/12/2020
					CT2750	Implantable cardiac electrical stimulators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	11/09/2020	10/12/2020
							F19	Surgical Intervention	A070504	Premature Discharge of Battery	11/09/2020	10/12/2020
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	11/09/2020	10/12/2020
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	11/09/2020	10/12/2020
							F19	Surgical Intervention	A070504	Premature Discharge of Battery	11/09/2020	10/12/2020

							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	11/09/2020	10/12/2020
					CT2985	Active	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	11/09/2020	10/12/2020
							F19	Surgical Intervention	A070504	Premature Discharge of Battery	11/09/2020	10/12/2020
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	11/09/2020	10/12/2020
					CT3001	Implantable metal	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	11/09/2020	10/12/2020
							F19	Surgical Intervention	A070504	Premature Discharge of Battery	11/09/2020	10/12/2020
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	11/09/2020	10/12/2020
					CT3002	Implantable synthetic polymer	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	11/09/2020	10/12/2020
							F19	Surgical Intervention	A070504	Premature Discharge of Battery	11/09/2020	10/12/2020
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	11/09/2020	10/12/2020
					CT3006	Implantable inorganic material	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	11/09/2020	10/12/2020
							F19	Surgical Intervention	A070504	Premature Discharge of Battery	11/09/2020	10/12/2020
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	11/09/2020	10/12/2020
					CT314	Electrically powered	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	11/09/2020	10/12/2020
							F19	Surgical Intervention	A070504	Premature Discharge of Battery	11/09/2020	10/12/2020
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	11/09/2020	10/12/2020
					CT321	Long-term surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	11/09/2020	10/12/2020
							F19	Surgical Intervention	A070504	Premature Discharge of Battery	11/09/2020	10/12/2020
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	11/09/2020	10/12/2020
					CT449	Generators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	11/09/2020	10/12/2020
							F19	Surgical Intervention	A070504	Premature Discharge of Battery	11/09/2020	10/12/2020
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	11/09/2020	10/12/2020
					CT981	Single use	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	11/09/2020	10/12/2020
							F19	Surgical Intervention	A070504	Premature Discharge of Battery	11/09/2020	10/12/2020
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	11/09/2020	10/12/2020
					CT983	Surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	11/09/2020	10/12/2020
							F19	Surgical Intervention	A070504	Premature Discharge of Battery	11/09/2020	10/12/2020
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	11/09/2020	10/12/2020
9975	2019	All other reportable incidents	36241	Pacemaker/defibrillator lead	CT1684	Cardiac pacemakers and associated devices	F1905	Device Revision or Replacement	A070905	High Sensing Threshold	11/10/2019	
9976	2019	All other reportable incidents	36241	Pacemaker/defibrillator lead	CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A0510	Retraction Problem	15/10/2019	
9977	2019	All other reportable incidents	47265	Dual-chamber implantable pacemaker, rate-responsive	CT141	Electrodes/Leads and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	11/12/2019	11/12/2019
					CT145	Cardiovascular devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	11/12/2019	11/12/2019
					CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	11/12/2019	11/12/2019
					CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	11/12/2019	11/12/2019
					CT233	Surgical	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	11/12/2019	11/12/2019
					CT2406	Implantable	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	11/12/2019	11/12/2019
					CT2407	Active implantable	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	11/12/2019	11/12/2019
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	11/12/2019	11/12/2019
					CT2727	Active implant devices/ancillaries	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	11/12/2019	11/12/2019
					CT2750	Implantable cardiac electrical stimulators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	11/12/2019	11/12/2019
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	11/12/2019	11/12/2019
					CT2985	Active	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	11/12/2019	11/12/2019
					CT3001	Implantable metal	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	11/12/2019	11/12/2019
					CT3002	Implantable synthetic polymer	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	11/12/2019	11/12/2019
					CT3006	Implantable inorganic material	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	11/12/2019	11/12/2019
					CT314	Electrically powered	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	11/12/2019	11/12/2019
					CT321	Long-term surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	11/12/2019	11/12/2019
					CT449	Generators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	11/12/2019	11/12/2019
					CT981	Single use	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	11/12/2019	11/12/2019
					CT983	Surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	11/12/2019	11/12/2019
9978	2020	All other reportable incidents	37265	Dual-chamber implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A072201	High Impedance		
							F26	No Health Consequences or Impact	A072201	High Impedance		
9979	2020	All other reportable incidents	47267	Single-chamber implantable pacemaker, rate-responsive	CT141	Electrodes/Leads and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F19	Surgical Intervention	A070504	Premature Discharge of Battery		
					CT145	Cardiovascular devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F19	Surgical Intervention	A070504	Premature Discharge of Battery		
					CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F19	Surgical Intervention	A070504	Premature Discharge of Battery		
					CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F19	Surgical Intervention	A070504	Premature Discharge of Battery		
					CT233	Surgical	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F19	Surgical Intervention	A070504	Premature Discharge of Battery		
					CT2406	Implantable	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F19	Surgical Intervention	A070504	Premature Discharge of Battery		
					CT2407	Active implantable	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F19	Surgical Intervention	A070504	Premature Discharge of Battery		
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F19	Surgical Intervention	A070504	Premature Discharge of Battery		
					CT2727	Active implant devices/ancillaries	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F19	Surgical Intervention	A070504	Premature Discharge of Battery		
					CT2750	Implantable cardiac electrical stimulators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F19	Surgical Intervention	A070504	Premature Discharge of Battery		
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F19	Surgical Intervention	A070504	Premature Discharge of Battery		
					CT2985	Active	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F19	Surgical Intervention	A070504	Premature Discharge of Battery		
					CT3001	Implantable metal	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F19	Surgical Intervention	A070504	Premature Discharge of Battery		
					CT3002	Implantable synthetic polymer	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F19	Surgical Intervention	A070504	Premature Discharge of Battery		
					CT3006	Implantable inorganic material	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F19	Surgical Intervention	A070504	Premature Discharge of Battery		
					CT314	Electrically powered	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F19	Surgical Intervention	A070504	Premature Discharge of Battery		
					CT321	Long-term surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F19	Surgical Intervention	A070504	Premature Discharge of Battery		
					CT449	Generators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		

							F19	Surgical Intervention	A070504	Premature Discharge of Battery		
					CT981	Single use	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F19	Surgical Intervention	A070504	Premature Discharge of Battery		
					CT983	Surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F19	Surgical Intervention	A070504	Premature Discharge of Battery		
9980	2020	All other reportable incidents	35852	Single-chamber implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	06/01/2021	06/01/2021
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	06/01/2021	06/01/2021
							F12	Serious Injury/ Illness/ Impairment	A070504	Premature Discharge of Battery	06/01/2021	06/01/2021
							F19	Surgical Intervention	A070504	Premature Discharge of Battery	06/01/2021	06/01/2021
9981	2020	All other reportable incidents	47270	Cardiac resynchronization therapy implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A0712	Pacing Problem		
							F26	No Health Consequences or Impact	A0712	Pacing Problem		
9982	2020	All other reportable incidents	47263	Cardiac resynchronization therapy implantable pacemaker	CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A0908	Incorrect, Inadequate or Imprecise Result or Readings	24/08/2020	24/08/2020
							F19	Surgical Intervention	A0908	Incorrect, Inadequate or Imprecise Result or Readings	24/08/2020	24/08/2020
					CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A0908	Incorrect, Inadequate or Imprecise Result or Readings	24/08/2020	24/08/2020
							F19	Surgical Intervention	A0908	Incorrect, Inadequate or Imprecise Result or Readings	24/08/2020	24/08/2020
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A0908	Incorrect, Inadequate or Imprecise Result or Readings	24/08/2020	24/08/2020
							F19	Surgical Intervention	A0908	Incorrect, Inadequate or Imprecise Result or Readings	24/08/2020	24/08/2020
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A0908	Incorrect, Inadequate or Imprecise Result or Readings	24/08/2020	24/08/2020
							F19	Surgical Intervention	A0908	Incorrect, Inadequate or Imprecise Result or Readings	24/08/2020	24/08/2020
9983	2020	All other reportable incidents	36241	Pacemaker/defibrillator lead	CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A072201	High Impedance		
							F26	No Health Consequences or Impact	A072201	High Impedance		
9984	2019	All other reportable incidents	47265	Dual-chamber implantable pacemaker, rate-responsive	CT141	Electrodes/Leads and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery		
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery		
					CT145	Cardiovascular devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery		
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery		
					CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery		
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery		
					CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery		
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery		
					CT233	Surgical	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery		
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery		
					CT2406	Implantable	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery		
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery		
					CT2407	Active implantable	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery		
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery		
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery		
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery		
					CT2727	Active implant devices/ancillaries	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery		
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery		
					CT2750	Implantable cardiac electrical stimulators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery		
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery		
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery		
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery		
					CT2985	Active	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery		
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery		
					CT3001	Implantable metal	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery		
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery		
					CT3002	Implantable synthetic polymer	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery		
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery		
					CT3006	Implantable inorganic material	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery		
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery		
					CT314	Electrically powered	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery		
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery		
					CT321	Long-term surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery		
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery		
					CT449	Generators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery		
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery		
					CT981	Single use	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery		
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery		
					CT983	Surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery		
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery		
9985	2019	All other reportable incidents	47267	Single-chamber implantable pacemaker, rate-	CT141	Electrodes/Leads and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070909	Oversensing		

				responsive	CT145	Cardiovascular devices	E2403	No Clinical Signs, Symptoms or Conditions	A070909	Oversensing		
					CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070909	Oversensing		
					CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A070909	Oversensing		
					CT233	Surgical	E2403	No Clinical Signs, Symptoms or Conditions	A070909	Oversensing		
					CT2406	Implantable	E2403	No Clinical Signs, Symptoms or Conditions	A070909	Oversensing		
					CT2407	Active implantable	E2403	No Clinical Signs, Symptoms or Conditions	A070909	Oversensing		
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A070909	Oversensing		
					CT2727	Active implant devices/ancillaries	E2403	No Clinical Signs, Symptoms or Conditions	A070909	Oversensing		
					CT2750	Implantable cardiac electrical stimulators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070909	Oversensing		
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A070909	Oversensing		
					CT2985	Active	E2403	No Clinical Signs, Symptoms or Conditions	A070909	Oversensing		
					CT3001	Implantable metal	E2403	No Clinical Signs, Symptoms or Conditions	A070909	Oversensing		
					CT3002	Implantable synthetic polymer	E2403	No Clinical Signs, Symptoms or Conditions	A070909	Oversensing		
					CT3006	Implantable inorganic material	E2403	No Clinical Signs, Symptoms or Conditions	A070909	Oversensing		
					CT314	Electrically powered	E2403	No Clinical Signs, Symptoms or Conditions	A070909	Oversensing		
					CT321	Long-term surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A070909	Oversensing		
					CT449	Generators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070909	Oversensing		
					CT981	Single use	E2403	No Clinical Signs, Symptoms or Conditions	A070909	Oversensing		
					CT983	Surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A070909	Oversensing		
9986	2019	All other reportable incidents	36241	Pacemaker/defibrillator lead	CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A05	Mechanical Problem	24/09/2020	24/09/2020
							F26	No Health Consequences or Impact	A05	Mechanical Problem	24/09/2020	24/09/2020
9987	2019	All other reportable incidents	37265	Dual-chamber implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A071302	Inappropriate/Inadequate Shock/Stimulation		
9988	2019	All other reportable incidents	47265	Dual-chamber implantable pacemaker, rate-responsive	CT141	Electrodes/Leads and associated devices	F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery		
					CT145	Cardiovascular devices	F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery		
					CT1684	Cardiac pacemakers and associated devices	F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery		
					CT1685	Cardiac pacemakers	F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery		
					CT233	Surgical	F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery		
					CT2406	Implantable	F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery		
					CT2407	Active implantable	F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery		
					CT257	Implantable cardiac pacemakers	F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery		
					CT2727	Active implant devices/ancillaries	F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery		
					CT2750	Implantable cardiac electrical stimulators and associated devices	F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery		
					CT2866	Implantable cardiac electrical stimulators	F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery		
					CT2985	Active	F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery		
					CT3001	Implantable metal	F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery		
					CT3002	Implantable synthetic polymer	F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery		
					CT3006	Implantable inorganic material	F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery		
					CT314	Electrically powered	F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery		
					CT321	Long-term surgical invasive	F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery		
					CT449	Generators and associated devices	F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery		
					CT981	Single use	F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery		
					CT983	Surgical invasive	F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery		
9989	2019	All other reportable incidents	47263	Cardiac resynchronization therapy implantable pacemaker	CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070909	Oversensing		
					CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A070909	Oversensing		
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A070909	Oversensing		
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A070909	Oversensing		
9990	2019	All other reportable incidents	47263	Cardiac resynchronization therapy implantable pacemaker	CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
					CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
9991	2020	All other reportable incidents	17882	Physiologic-monitoring defibrillation system	CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A0719	Unexpected Shutdown		
							F26	No Health Consequences or Impact	A0719	Unexpected Shutdown		
9992	2020	All other reportable incidents	35852	Single-chamber implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A070909	Oversensing		
									A071302	Inappropriate/Inadequate Shock/Stimulation		
							F26	No Health Consequences or Impact	A070909	Oversensing		
									A071302	Inappropriate/Inadequate Shock/Stimulation		
9993	2020	All other reportable incidents	37265	Dual-chamber implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A070909	Oversensing		
							F26	No Health Consequences or Impact	A070909	Oversensing		
9994	2020	All other reportable incidents	60789	Intracardiac pacemaker	CT130	Catheters and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	07/12/2020	07/12/2020
									A071102	Failure to Interrogate	07/12/2020	07/12/2020
							F1901	Additional Surgery	A070504	Premature Discharge of Battery	07/12/2020	07/12/2020
									A071102	Failure to Interrogate	07/12/2020	07/12/2020
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	07/12/2020	07/12/2020
									A071102	Failure to Interrogate	07/12/2020	07/12/2020
					CT145	Cardiovascular devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	07/12/2020	07/12/2020
									A071102	Failure to Interrogate	07/12/2020	07/12/2020
							F1901	Additional Surgery	A070504	Premature Discharge of Battery	07/12/2020	07/12/2020
									A071102	Failure to Interrogate	07/12/2020	07/12/2020
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	07/12/2020	07/12/2020
									A071102	Failure to Interrogate	07/12/2020	07/12/2020
					CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	07/12/2020	07/12/2020
									A071102	Failure to Interrogate	07/12/2020	07/12/2020
							F1901	Additional Surgery	A070504	Premature Discharge of Battery	07/12/2020	07/12/2020
									A071102	Failure to Interrogate	07/12/2020	07/12/2020
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	07/12/2020	07/12/2020
									A071102	Failure to Interrogate	07/12/2020	07/12/2020
					CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	07/12/2020	07/12/2020
									A071102	Failure to Interrogate	07/12/2020	07/12/2020
							F1901	Additional Surgery	A070504	Premature Discharge of Battery	07/12/2020	07/12/2020
									A071102	Failure to Interrogate	07/12/2020	07/12/2020
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	07/12/2020	07/12/2020
									A071102	Failure to Interrogate	07/12/2020	07/12/2020
					CT233	Surgical	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	07/12/2020	07/12/2020
									A071102	Failure to Interrogate	07/12/2020	07/12/2020
							F1901	Additional Surgery	A070504	Premature Discharge of Battery	07/12/2020	07/12/2020
									A071102	Failure to Interrogate	07/12/2020	07/12/2020

						F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	07/12/2020	07/12/2020
								A071102	Failure to Interrogate	07/12/2020	07/12/2020
					CT2406	Implantable	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	07/12/2020	07/12/2020
									A071102	Failure to Interrogate	07/12/2020	07/12/2020
							F1901	Additional Surgery	A070504	Premature Discharge of Battery	07/12/2020	07/12/2020
									A071102	Failure to Interrogate	07/12/2020	07/12/2020
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	07/12/2020	07/12/2020
									A071102	Failure to Interrogate	07/12/2020	07/12/2020
					CT2407	Active implantable	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	07/12/2020	07/12/2020
									A071102	Failure to Interrogate	07/12/2020	07/12/2020
							F1901	Additional Surgery	A070504	Premature Discharge of Battery	07/12/2020	07/12/2020
									A071102	Failure to Interrogate	07/12/2020	07/12/2020
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	07/12/2020	07/12/2020
									A071102	Failure to Interrogate	07/12/2020	07/12/2020
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	07/12/2020	07/12/2020
									A071102	Failure to Interrogate	07/12/2020	07/12/2020
							F1901	Additional Surgery	A070504	Premature Discharge of Battery	07/12/2020	07/12/2020
									A071102	Failure to Interrogate	07/12/2020	07/12/2020
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	07/12/2020	07/12/2020
									A071102	Failure to Interrogate	07/12/2020	07/12/2020
					CT2727	Active implant devices/ancillaries	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	07/12/2020	07/12/2020
									A071102	Failure to Interrogate	07/12/2020	07/12/2020
							F1901	Additional Surgery	A070504	Premature Discharge of Battery	07/12/2020	07/12/2020
									A071102	Failure to Interrogate	07/12/2020	07/12/2020
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	07/12/2020	07/12/2020
									A071102	Failure to Interrogate	07/12/2020	07/12/2020
					CT2750	Implantable cardiac electrical stimulators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	07/12/2020	07/12/2020
									A071102	Failure to Interrogate	07/12/2020	07/12/2020
							F1901	Additional Surgery	A070504	Premature Discharge of Battery	07/12/2020	07/12/2020
									A071102	Failure to Interrogate	07/12/2020	07/12/2020
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	07/12/2020	07/12/2020
									A071102	Failure to Interrogate	07/12/2020	07/12/2020
					CT2818	Catheterization/Cannulation/Fluid path devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	07/12/2020	07/12/2020
									A071102	Failure to Interrogate	07/12/2020	07/12/2020
							F1901	Additional Surgery	A070504	Premature Discharge of Battery	07/12/2020	07/12/2020
									A071102	Failure to Interrogate	07/12/2020	07/12/2020
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	07/12/2020	07/12/2020
									A071102	Failure to Interrogate	07/12/2020	07/12/2020
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	07/12/2020	07/12/2020
									A071102	Failure to Interrogate	07/12/2020	07/12/2020
							F1901	Additional Surgery	A070504	Premature Discharge of Battery	07/12/2020	07/12/2020
									A071102	Failure to Interrogate	07/12/2020	07/12/2020
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	07/12/2020	07/12/2020
									A071102	Failure to Interrogate	07/12/2020	07/12/2020
					CT2985	Active	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	07/12/2020	07/12/2020
									A071102	Failure to Interrogate	07/12/2020	07/12/2020
							F1901	Additional Surgery	A070504	Premature Discharge of Battery	07/12/2020	07/12/2020
									A071102	Failure to Interrogate	07/12/2020	07/12/2020
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	07/12/2020	07/12/2020
									A071102	Failure to Interrogate	07/12/2020	07/12/2020
					CT3001	Implantable metal	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	07/12/2020	07/12/2020
									A071102	Failure to Interrogate	07/12/2020	07/12/2020
							F1901	Additional Surgery	A070504	Premature Discharge of Battery	07/12/2020	07/12/2020
									A071102	Failure to Interrogate	07/12/2020	07/12/2020
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	07/12/2020	07/12/2020
									A071102	Failure to Interrogate	07/12/2020	07/12/2020
					CT3002	Implantable synthetic polymer	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	07/12/2020	07/12/2020
									A071102	Failure to Interrogate	07/12/2020	07/12/2020
							F1901	Additional Surgery	A070504	Premature Discharge of Battery	07/12/2020	07/12/2020
									A071102	Failure to Interrogate	07/12/2020	07/12/2020
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	07/12/2020	07/12/2020
									A071102	Failure to Interrogate	07/12/2020	07/12/2020
					CT3006	Implantable inorganic material	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	07/12/2020	07/12/2020
									A071102	Failure to Interrogate	07/12/2020	07/12/2020
							F1901	Additional Surgery	A070504	Premature Discharge of Battery	07/12/2020	07/12/2020
									A071102	Failure to Interrogate	07/12/2020	07/12/2020
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	07/12/2020	07/12/2020
									A071102	Failure to Interrogate	07/12/2020	07/12/2020
					CT314	Electrically powered	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	07/12/2020	07/12/2020
									A071102	Failure to Interrogate	07/12/2020	07/12/2020
							F1901	Additional Surgery	A070504	Premature Discharge of Battery	07/12/2020	07/12/2020
									A071102	Failure to Interrogate	07/12/2020	07/12/2020
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	07/12/2020	07/12/2020
									A071102	Failure to Interrogate	07/12/2020	07/12/2020
					CT321	Long-term surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	07/12/2020	07/12/2020
									A071102	Failure to Interrogate	07/12/2020	07/12/2020
							F1901	Additional Surgery	A070504	Premature Discharge of Battery	07/12/2020	07/12/2020
									A071102	Failure to Interrogate	07/12/2020	07/12/2020
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	07/12/2020	07/12/2020
									A071102	Failure to Interrogate	07/12/2020	07/12/2020
					CT981	Single use	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	07/12/2020	07/12/2020
									A071102	Failure to Interrogate	07/12/2020	07/12/2020
							F1901	Additional Surgery	A070504	Premature Discharge of Battery	07/12/2020	07/12/2020
									A071102	Failure to Interrogate	07/12/2020	07/12/2020
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	07/12/2020	07/12/2020
									A071102	Failure to Interrogate	07/12/2020	07/12/2020
					CT983	Surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	07/12/2020	07/12/2020
									A071102	Failure to Interrogate	07/12/2020	07/12/2020

							F1901	Additional Surgery	A070504	Premature Discharge of Battery	07/12/2020	07/12/2020
									A071102	Failure to Interrogate	07/12/2020	07/12/2020
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	07/12/2020	07/12/2020
									A071102	Failure to Interrogate	07/12/2020	07/12/2020
9995	2020	All other reportable incidents	35852	Single-chamber implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A0908	Incorrect, Inadequate or Imprecise Result or Readings		
									A1107	Data Problem		
							F24	Insufficient Health Impact Information	A0908	Incorrect, Inadequate or Imprecise Result or Readings		
									A1107	Data Problem		
9996	2020	All other reportable incidents	35852	Single-chamber implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A070909	Oversensing		
							F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A070909	Oversensing		
							F26	No Health Consequences or Impact	A070909	Oversensing		
9997	2019	All other reportable incidents	12913	Pacemaker, implantable	CT145	Cardiovascular devices	E0605	Cardiac Tamponade	A07	Electrical /Electronic Property Problem		
							F1905	Device Revision or Replacement	A07	Electrical /Electronic Property Problem		
					CT1684	Cardiac pacemakers and associated devices	E0605	Cardiac Tamponade	A07	Electrical /Electronic Property Problem		
							F1905	Device Revision or Replacement	A07	Electrical /Electronic Property Problem		
					CT1685	Cardiac pacemakers	E0605	Cardiac Tamponade	A07	Electrical /Electronic Property Problem		
							F1905	Device Revision or Replacement	A07	Electrical /Electronic Property Problem		
					CT2406	Implantable	E0605	Cardiac Tamponade	A07	Electrical /Electronic Property Problem		
							F1905	Device Revision or Replacement	A07	Electrical /Electronic Property Problem		
					CT2407	Active implantable	E0605	Cardiac Tamponade	A07	Electrical /Electronic Property Problem		
							F1905	Device Revision or Replacement	A07	Electrical /Electronic Property Problem		
					CT257	Implantable cardiac pacemakers	E0605	Cardiac Tamponade	A07	Electrical /Electronic Property Problem		
							F1905	Device Revision or Replacement	A07	Electrical /Electronic Property Problem		
					CT2727	Active implant devices/ancillaries	E0605	Cardiac Tamponade	A07	Electrical /Electronic Property Problem		
							F1905	Device Revision or Replacement	A07	Electrical /Electronic Property Problem		
					CT2750	Implantable cardiac electrical stimulators and associated devices	E0605	Cardiac Tamponade	A07	Electrical /Electronic Property Problem		
							F1905	Device Revision or Replacement	A07	Electrical /Electronic Property Problem		
					CT2866	Implantable cardiac electrical stimulators	E0605	Cardiac Tamponade	A07	Electrical /Electronic Property Problem		
							F1905	Device Revision or Replacement	A07	Electrical /Electronic Property Problem		
					CT2985	Active	E0605	Cardiac Tamponade	A07	Electrical /Electronic Property Problem		
							F1905	Device Revision or Replacement	A07	Electrical /Electronic Property Problem		
					CT314	Electrically powered	E0605	Cardiac Tamponade	A07	Electrical /Electronic Property Problem		
							F1905	Device Revision or Replacement	A07	Electrical /Electronic Property Problem		
9998	2019	All other reportable incidents	35852	Single-chamber implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A072201	High Impedance		
9999	2019	All other reportable incidents	35852	Single-chamber implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A071302	Inappropriate/Inadequate Shock/Stimulation		
10000	2019	All other reportable incidents	37265	Dual-chamber implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A072201	High Impedance		
10001	2019	All other reportable incidents	35223	Endocardial/interventricular septal pacing lead	CT1684	Cardiac pacemakers and associated devices	F1905	Device Revision or Replacement	A070909	Oversensing	13/09/2019	
10002	2020	All other reportable incidents	47270	Cardiac resynchronization therapy implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	16/09/2020	16/09/2020
							F1901	Additional Surgery	A070504	Premature Discharge of Battery	16/09/2020	16/09/2020
10003	2020	All other reportable incidents	17882	Physiologic-monitoring defibrillation system	CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A07	Electrical /Electronic Property Problem		
							F26	No Health Consequences or Impact	A07	Electrical /Electronic Property Problem		
10004	2020	All other reportable incidents	17882	Physiologic-monitoring defibrillation system	CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070803	Failure to Power Up		
							F27	Problem Identified During Non-Clinical Procedure	A070803	Failure to Power Up		
10005	2020	All other reportable incidents	17882	Physiologic-monitoring defibrillation system	CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A0719	Unexpected Shutdown		
							F26	No Health Consequences or Impact	A0719	Unexpected Shutdown		
10006	2020	All other reportable incidents	35852	Single-chamber implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A070909	Oversensing		
							F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A070909	Oversensing		
							F26	No Health Consequences or Impact	A070909	Oversensing		
10007	2020	All other reportable incidents	36241	Pacemaker/defibrillator lead	CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A072201	High Impedance		
							F24	Insufficient Health Impact Information	A072201	High Impedance		
10008	2020	All other reportable incidents	47265	Dual-chamber implantable pacemaker, rate-responsive	CT141	Electrodes/Leads and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A0908	Incorrect, Inadequate or Imprecise Result or Readings	01/09/2020	01/09/2020
							F08	Hospitalization or Prolonged Hospitalization	A0908	Incorrect, Inadequate or Imprecise Result or Readings	01/09/2020	01/09/2020
							F1905	Device Revision or Replacement	A0908	Incorrect, Inadequate or Imprecise Result or Readings	01/09/2020	01/09/2020
					CT145	Cardiovascular devices	E2403	No Clinical Signs, Symptoms or Conditions	A0908	Incorrect, Inadequate or Imprecise Result or Readings	01/09/2020	01/09/2020
							F08	Hospitalization or Prolonged Hospitalization	A0908	Incorrect, Inadequate or Imprecise Result or Readings	01/09/2020	01/09/2020
							F1905	Device Revision or Replacement	A0908	Incorrect, Inadequate or Imprecise Result or Readings	01/09/2020	01/09/2020
					CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A0908	Incorrect, Inadequate or Imprecise Result or Readings	01/09/2020	01/09/2020
							F08	Hospitalization or Prolonged Hospitalization	A0908	Incorrect, Inadequate or Imprecise Result or Readings	01/09/2020	01/09/2020
							F1905	Device Revision or Replacement	A0908	Incorrect, Inadequate or Imprecise Result or Readings	01/09/2020	01/09/2020
					CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A0908	Incorrect, Inadequate or Imprecise Result or Readings	01/09/2020	01/09/2020
							F08	Hospitalization or Prolonged Hospitalization	A0908	Incorrect, Inadequate or Imprecise Result or Readings	01/09/2020	01/09/2020
							F1905	Device Revision or Replacement	A0908	Incorrect, Inadequate or Imprecise Result or Readings	01/09/2020	01/09/2020
					CT233	Surgical	E2403	No Clinical Signs, Symptoms or Conditions	A0908	Incorrect, Inadequate or Imprecise Result or Readings	01/09/2020	01/09/2020
							F08	Hospitalization or Prolonged Hospitalization	A0908	Incorrect, Inadequate or Imprecise Result or Readings	01/09/2020	01/09/2020
							F1905	Device Revision or Replacement	A0908	Incorrect, Inadequate or Imprecise Result or Readings	01/09/2020	01/09/2020
					CT2406	Implantable	E2403	No Clinical Signs, Symptoms or Conditions	A0908	Incorrect, Inadequate or Imprecise Result or Readings	01/09/2020	01/09/2020
							F08	Hospitalization or Prolonged Hospitalization	A0908	Incorrect, Inadequate or Imprecise Result or Readings	01/09/2020	01/09/2020
							F1905	Device Revision or Replacement	A0908	Incorrect, Inadequate or Imprecise Result or Readings	01/09/2020	01/09/2020
					CT2407	Active implantable	E2403	No Clinical Signs, Symptoms or Conditions	A0908	Incorrect, Inadequate or Imprecise Result or Readings	01/09/2020	01/09/2020
							F08	Hospitalization or Prolonged Hospitalization	A0908	Incorrect, Inadequate or Imprecise Result or Readings	01/09/2020	01/09/2020

							F1905	Device Revision or Replacement	A0908	Incorrect, Inadequate or Imprecise Result or Readings	01/09/2020	01/09/2020
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A0908	Incorrect, Inadequate or Imprecise Result or Readings	01/09/2020	01/09/2020
							F08	Hospitalization or Prolonged Hospitalization	A0908	Incorrect, Inadequate or Imprecise Result or Readings	01/09/2020	01/09/2020
							F1905	Device Revision or Replacement	A0908	Incorrect, Inadequate or Imprecise Result or Readings	01/09/2020	01/09/2020
					CT2727	Active implant devices/ancillaries	E2403	No Clinical Signs, Symptoms or Conditions	A0908	Incorrect, Inadequate or Imprecise Result or Readings	01/09/2020	01/09/2020
							F08	Hospitalization or Prolonged Hospitalization	A0908	Incorrect, Inadequate or Imprecise Result or Readings	01/09/2020	01/09/2020
							F1905	Device Revision or Replacement	A0908	Incorrect, Inadequate or Imprecise Result or Readings	01/09/2020	01/09/2020
					CT2750	Implantable cardiac electrical stimulators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A0908	Incorrect, Inadequate or Imprecise Result or Readings	01/09/2020	01/09/2020
							F08	Hospitalization or Prolonged Hospitalization	A0908	Incorrect, Inadequate or Imprecise Result or Readings	01/09/2020	01/09/2020
							F1905	Device Revision or Replacement	A0908	Incorrect, Inadequate or Imprecise Result or Readings	01/09/2020	01/09/2020
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A0908	Incorrect, Inadequate or Imprecise Result or Readings	01/09/2020	01/09/2020
							F08	Hospitalization or Prolonged Hospitalization	A0908	Incorrect, Inadequate or Imprecise Result or Readings	01/09/2020	01/09/2020
							F1905	Device Revision or Replacement	A0908	Incorrect, Inadequate or Imprecise Result or Readings	01/09/2020	01/09/2020
					CT2985	Active	E2403	No Clinical Signs, Symptoms or Conditions	A0908	Incorrect, Inadequate or Imprecise Result or Readings	01/09/2020	01/09/2020
							F08	Hospitalization or Prolonged Hospitalization	A0908	Incorrect, Inadequate or Imprecise Result or Readings	01/09/2020	01/09/2020
							F1905	Device Revision or Replacement	A0908	Incorrect, Inadequate or Imprecise Result or Readings	01/09/2020	01/09/2020
					CT3001	Implantable metal	E2403	No Clinical Signs, Symptoms or Conditions	A0908	Incorrect, Inadequate or Imprecise Result or Readings	01/09/2020	01/09/2020
							F08	Hospitalization or Prolonged Hospitalization	A0908	Incorrect, Inadequate or Imprecise Result or Readings	01/09/2020	01/09/2020
							F1905	Device Revision or Replacement	A0908	Incorrect, Inadequate or Imprecise Result or Readings	01/09/2020	01/09/2020
					CT3002	Implantable synthetic polymer	E2403	No Clinical Signs, Symptoms or Conditions	A0908	Incorrect, Inadequate or Imprecise Result or Readings	01/09/2020	01/09/2020
							F08	Hospitalization or Prolonged Hospitalization	A0908	Incorrect, Inadequate or Imprecise Result or Readings	01/09/2020	01/09/2020
							F1905	Device Revision or Replacement	A0908	Incorrect, Inadequate or Imprecise Result or Readings	01/09/2020	01/09/2020
					CT3006	Implantable inorganic material	E2403	No Clinical Signs, Symptoms or Conditions	A0908	Incorrect, Inadequate or Imprecise Result or Readings	01/09/2020	01/09/2020
							F08	Hospitalization or Prolonged Hospitalization	A0908	Incorrect, Inadequate or Imprecise Result or Readings	01/09/2020	01/09/2020
							F1905	Device Revision or Replacement	A0908	Incorrect, Inadequate or Imprecise Result or Readings	01/09/2020	01/09/2020
					CT314	Electrically powered	E2403	No Clinical Signs, Symptoms or Conditions	A0908	Incorrect, Inadequate or Imprecise Result or Readings	01/09/2020	01/09/2020
							F08	Hospitalization or Prolonged Hospitalization	A0908	Incorrect, Inadequate or Imprecise Result or Readings	01/09/2020	01/09/2020
							F1905	Device Revision or Replacement	A0908	Incorrect, Inadequate or Imprecise Result or Readings	01/09/2020	01/09/2020
					CT321	Long-term surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A0908	Incorrect, Inadequate or Imprecise Result or Readings	01/09/2020	01/09/2020
							F08	Hospitalization or Prolonged Hospitalization	A0908	Incorrect, Inadequate or Imprecise Result or Readings	01/09/2020	01/09/2020
							F1905	Device Revision or Replacement	A0908	Incorrect, Inadequate or Imprecise Result or Readings	01/09/2020	01/09/2020
					CT449	Generators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A0908	Incorrect, Inadequate or Imprecise Result or Readings	01/09/2020	01/09/2020
							F08	Hospitalization or Prolonged Hospitalization	A0908	Incorrect, Inadequate or Imprecise Result or Readings	01/09/2020	01/09/2020
							F1905	Device Revision or Replacement	A0908	Incorrect, Inadequate or Imprecise Result or Readings	01/09/2020	01/09/2020
					CT981	Single use	E2403	No Clinical Signs, Symptoms or Conditions	A0908	Incorrect, Inadequate or Imprecise Result or Readings	01/09/2020	01/09/2020
							F08	Hospitalization or Prolonged Hospitalization	A0908	Incorrect, Inadequate or Imprecise Result or Readings	01/09/2020	01/09/2020
							F1905	Device Revision or Replacement	A0908	Incorrect, Inadequate or Imprecise Result or Readings	01/09/2020	01/09/2020
					CT983	Surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A0908	Incorrect, Inadequate or Imprecise Result or Readings	01/09/2020	01/09/2020
							F08	Hospitalization or Prolonged Hospitalization	A0908	Incorrect, Inadequate or Imprecise Result or Readings	01/09/2020	01/09/2020
							F1905	Device Revision or Replacement	A0908	Incorrect, Inadequate or Imprecise Result or Readings	01/09/2020	01/09/2020
10009	2019	All other reportable incidents	36241	Pacemaker/defibrillator lead	CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070101	Failure to Capture		
10010	2019	All other reportable incidents	36241	Pacemaker/defibrillator lead	CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A072201	High Impedance	26/09/2019	26/09/2019
10011	2019	All other reportable incidents	47265	Dual-chamber implantable pacemaker, rate-responsive	CT141	Electrodes/Leads and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	05/12/2019	05/12/2019
					CT145	Cardiovascular devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	05/12/2019	05/12/2019
					CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	05/12/2019	05/12/2019
					CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	05/12/2019	05/12/2019
					CT233	Surgical	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	05/12/2019	05/12/2019
					CT2406	Implantable	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	05/12/2019	05/12/2019
					CT2407	Active implantable	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	05/12/2019	05/12/2019
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	05/12/2019	05/12/2019
					CT2727	Active implant devices/ancillaries	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	05/12/2019	05/12/2019
					CT2750	Implantable cardiac electrical stimulators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	05/12/2019	05/12/2019
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	05/12/2019	05/12/2019
					CT2985	Active	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	05/12/2019	05/12/2019
					CT3001	Implantable metal	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	05/12/2019	05/12/2019
					CT3002	Implantable synthetic polymer	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	05/12/2019	05/12/2019

					CT3006	Implantable inorganic material	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	05/12/2019	05/12/2019
					CT314	Electrically powered	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	05/12/2019	05/12/2019
					CT321	Long-term surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	05/12/2019	05/12/2019
					CT449	Generators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	05/12/2019	05/12/2019
					CT981	Single use	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	05/12/2019	05/12/2019
					CT983	Surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	05/12/2019	05/12/2019
10012	2019	All other reportable incidents	37265	Dual-chamber implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery		
10013	2019	All other reportable incidents	37265	Dual-chamber implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A070909	Oversensing		
10014	2019	All other reportable incidents	47267	Single-chamber implantable pacemaker, rate-responsive	CT141	Electrodes/Leads and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A072201	High Impedance		
					CT145	Cardiovascular devices	E2403	No Clinical Signs, Symptoms or Conditions	A072201	High Impedance		
					CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A072201	High Impedance		
					CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A072201	High Impedance		
					CT233	Surgical	E2403	No Clinical Signs, Symptoms or Conditions	A072201	High Impedance		
					CT2406	Implantable	E2403	No Clinical Signs, Symptoms or Conditions	A072201	High Impedance		
					CT2407	Active implantable	E2403	No Clinical Signs, Symptoms or Conditions	A072201	High Impedance		
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A072201	High Impedance		
					CT2727	Active implant devices/ancillaries	E2403	No Clinical Signs, Symptoms or Conditions	A072201	High Impedance		
					CT2750	Implantable cardiac electrical stimulators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A072201	High Impedance		
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A072201	High Impedance		
					CT2985	Active	E2403	No Clinical Signs, Symptoms or Conditions	A072201	High Impedance		
					CT3001	Implantable metal	E2403	No Clinical Signs, Symptoms or Conditions	A072201	High Impedance		
					CT3002	Implantable synthetic polymer	E2403	No Clinical Signs, Symptoms or Conditions	A072201	High Impedance		
					CT3006	Implantable inorganic material	E2403	No Clinical Signs, Symptoms or Conditions	A072201	High Impedance		
					CT314	Electrically powered	E2403	No Clinical Signs, Symptoms or Conditions	A072201	High Impedance		
					CT321	Long-term surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A072201	High Impedance		
					CT449	Generators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A072201	High Impedance		
					CT981	Single use	E2403	No Clinical Signs, Symptoms or Conditions	A072201	High Impedance		
					CT983	Surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A072201	High Impedance		
10015	2019	All other reportable incidents	35852	Single-chamber implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	F05	Delay to Treatment/ Therapy	A070909	Oversensing		
10016	2019	All other reportable incidents	47263	Cardiac resynchronization therapy implantable pacemaker	CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	01/06/2020	01/06/2020
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	01/06/2020	01/06/2020
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	01/06/2020	01/06/2020
					CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	01/06/2020	01/06/2020
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	01/06/2020	01/06/2020
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	01/06/2020	01/06/2020
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	01/06/2020	01/06/2020
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	01/06/2020	01/06/2020
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	01/06/2020	01/06/2020
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	01/06/2020	01/06/2020
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	01/06/2020	01/06/2020
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	01/06/2020	01/06/2020
10017	2020	All other reportable incidents	35852	Single-chamber implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A072201	High Impedance		
									A12	Connection Problem		
							F24	Insufficient Health Impact Information	A072201	High Impedance		
									A12	Connection Problem		
10018	2020	All other reportable incidents	47265	Dual-chamber implantable pacemaker, rate-responsive	CT141	Electrodes/Leads and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	06/10/2020	06/10/2020
							F19	Surgical Intervention	A070504	Premature Discharge of Battery	06/10/2020	06/10/2020
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	06/10/2020	06/10/2020
					CT145	Cardiovascular devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	06/10/2020	06/10/2020
							F19	Surgical Intervention	A070504	Premature Discharge of Battery	06/10/2020	06/10/2020
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	06/10/2020	06/10/2020
					CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	06/10/2020	06/10/2020
							F19	Surgical Intervention	A070504	Premature Discharge of Battery	06/10/2020	06/10/2020
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	06/10/2020	06/10/2020
					CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	06/10/2020	06/10/2020
							F19	Surgical Intervention	A070504	Premature Discharge of Battery	06/10/2020	06/10/2020
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	06/10/2020	06/10/2020
					CT233	Surgical	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	06/10/2020	06/10/2020
							F19	Surgical Intervention	A070504	Premature Discharge of Battery	06/10/2020	06/10/2020
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	06/10/2020	06/10/2020
					CT2406	Implantable	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	06/10/2020	06/10/2020
							F19	Surgical Intervention	A070504	Premature Discharge of Battery	06/10/2020	06/10/2020
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	06/10/2020	06/10/2020
					CT2407	Active implantable	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	06/10/2020	06/10/2020
							F19	Surgical Intervention	A070504	Premature Discharge of Battery	06/10/2020	06/10/2020
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	06/10/2020	06/10/2020
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	06/10/2020	06/10/2020
							F19	Surgical Intervention	A070504	Premature Discharge of Battery	06/10/2020	06/10/2020
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	06/10/2020	06/10/2020
					CT2727	Active implant devices/ancillaries	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	06/10/2020	06/10/2020
							F19	Surgical Intervention	A070504	Premature Discharge of Battery	06/10/2020	06/10/2020
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	06/10/2020	06/10/2020
					CT2750	Implantable cardiac electrical stimulators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	06/10/2020	06/10/2020
							F19	Surgical Intervention	A070504	Premature Discharge of Battery	06/10/2020	06/10/2020
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	06/10/2020	06/10/2020
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	06/10/2020	06/10/2020
							F19	Surgical Intervention	A070504	Premature Discharge of Battery	06/10/2020	06/10/2020
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	06/10/2020	06/10/2020
					CT2985	Active	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	06/10/2020	06/10/2020
							F19	Surgical Intervention	A070504	Premature Discharge of Battery	06/10/2020	06/10/2020
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	06/10/2020	06/10/2020
					CT3001	Implantable metal	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	06/10/2020	06/10/2020
							F19	Surgical Intervention	A070504	Premature Discharge of Battery	06/10/2020	06/10/2020
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	06/10/2020	06/10/2020
					CT3002	Implantable synthetic polymer	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	06/10/2020	06/10/2020
							F19	Surgical Intervention	A070504	Premature Discharge of Battery	06/10/2020	06/10/2020
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	06/10/2020	06/10/2020

					CT3006	Implantable inorganic material	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	06/10/2020	06/10/2020
							F19	Surgical Intervention	A070504	Premature Discharge of Battery	06/10/2020	06/10/2020
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	06/10/2020	06/10/2020
					CT314	Electrically powered	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	06/10/2020	06/10/2020
							F19	Surgical Intervention	A070504	Premature Discharge of Battery	06/10/2020	06/10/2020
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	06/10/2020	06/10/2020
					CT321	Long-term surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	06/10/2020	06/10/2020
							F19	Surgical Intervention	A070504	Premature Discharge of Battery	06/10/2020	06/10/2020
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	06/10/2020	06/10/2020
					CT449	Generators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	06/10/2020	06/10/2020
							F19	Surgical Intervention	A070504	Premature Discharge of Battery	06/10/2020	06/10/2020
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	06/10/2020	06/10/2020
					CT981	Single use	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	06/10/2020	06/10/2020
							F19	Surgical Intervention	A070504	Premature Discharge of Battery	06/10/2020	06/10/2020
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	06/10/2020	06/10/2020
					CT983	Surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	06/10/2020	06/10/2020
							F19	Surgical Intervention	A070504	Premature Discharge of Battery	06/10/2020	06/10/2020
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	06/10/2020	06/10/2020
10019	2020	All other reportable incidents	36241	Pacemaker/defibrillator lead	CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A05	Mechanical Problem	04/09/2020	04/09/2020
									A051201	Device Dislodged or Dislocated	04/09/2020	04/09/2020
									A2304	Off-Label Use	04/09/2020	04/09/2020
							F19	Surgical Intervention	A05	Mechanical Problem	04/09/2020	04/09/2020
									A051201	Device Dislodged or Dislocated	04/09/2020	04/09/2020
									A2304	Off-Label Use	04/09/2020	04/09/2020
							F1905	Device Revision or Replacement	A05	Mechanical Problem	04/09/2020	04/09/2020
									A051201	Device Dislodged or Dislocated	04/09/2020	04/09/2020
									A2304	Off-Label Use	04/09/2020	04/09/2020
10020	2020	Death	37265	Dual-chamber implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	11/09/2020	11/09/2020
							F02	Death	A070504	Premature Discharge of Battery	11/09/2020	11/09/2020
10021	2020	All other reportable incidents	47265	Dual-chamber implantable pacemaker, rate-responsive	CT141	Electrodes/Leads and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem		
							F26	No Health Consequences or Impact	A0705	Battery Problem		
					CT145	Cardiovascular devices	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem		
							F26	No Health Consequences or Impact	A0705	Battery Problem		
					CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem		
							F26	No Health Consequences or Impact	A0705	Battery Problem		
					CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem		
							F26	No Health Consequences or Impact	A0705	Battery Problem		
					CT233	Surgical	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem		
							F26	No Health Consequences or Impact	A0705	Battery Problem		
					CT2406	Implantable	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem		
							F26	No Health Consequences or Impact	A0705	Battery Problem		
					CT2407	Active implantable	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem		
							F26	No Health Consequences or Impact	A0705	Battery Problem		
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem		
							F26	No Health Consequences or Impact	A0705	Battery Problem		
					CT2727	Active implant devices/ancillaries	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem		
							F26	No Health Consequences or Impact	A0705	Battery Problem		
					CT2750	Implantable cardiac electrical stimulators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem		
							F26	No Health Consequences or Impact	A0705	Battery Problem		
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem		
							F26	No Health Consequences or Impact	A0705	Battery Problem		
					CT2985	Active	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem		
							F26	No Health Consequences or Impact	A0705	Battery Problem		
					CT3001	Implantable metal	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem		
							F26	No Health Consequences or Impact	A0705	Battery Problem		
					CT3002	Implantable synthetic polymer	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem		
							F26	No Health Consequences or Impact	A0705	Battery Problem		
					CT3006	Implantable inorganic material	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem		
							F26	No Health Consequences or Impact	A0705	Battery Problem		
					CT314	Electrically powered	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem		
							F26	No Health Consequences or Impact	A0705	Battery Problem		
					CT321	Long-term surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem		
							F26	No Health Consequences or Impact	A0705	Battery Problem		
					CT449	Generators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem		
							F26	No Health Consequences or Impact	A0705	Battery Problem		
					CT981	Single use	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem		
							F26	No Health Consequences or Impact	A0705	Battery Problem		
					CT983	Surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem		
							F26	No Health Consequences or Impact	A0705	Battery Problem		
10022	2020	All other reportable incidents	36241	Pacemaker/defibrillator lead	CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070101	Failure to Capture		
									A0722	Impedance Problem		
							F26	No Health Consequences or Impact	A070101	Failure to Capture		
									A0722	Impedance Problem		
10023	2020	All other reportable incidents	36241	Pacemaker/defibrillator lead	CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A051201	Device Dislodged or Dislocated		
									A0722	Impedance Problem		
							F26	No Health Consequences or Impact	A051201	Device Dislodged or Dislocated		
									A0722	Impedance Problem		
10024	2020	All other reportable incidents	47263	Cardiac resynchronization therapy implantable pacemaker	CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070101	Failure to Capture		
							F26	No Health Consequences or Impact	A070101	Failure to Capture		
					CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A070101	Failure to Capture		
							F26	No Health Consequences or Impact	A070101	Failure to Capture		
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A070101	Failure to Capture		
							F26	No Health Consequences or Impact	A070101	Failure to Capture		
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A070101	Failure to Capture		
							F26	No Health Consequences or Impact	A070101	Failure to Capture		
10025	2019	All other reportable incidents	37265	Dual-chamber implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	11/10/2019	
10026	2019	All other reportable incidents	36241	Pacemaker/defibrillator lead	CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070909	Oversensing		
10027	2019	All other reportable incidents	35852	Single-chamber implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A070909	Oversensing		

10028	2019	All other reportable incidents	36241	Pacemaker/defibrillator lead	CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070910	Undersensing		
10029	2020	All other reportable incidents	47267	Single-chamber implantable pacemaker, rate-responsive	CT141	Electrodes/Leads and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A071102	Failure to Interrogate		
							F23	Unexpected Medical Intervention	A071102	Failure to Interrogate		
					CT145	Cardiovascular devices	E2403	No Clinical Signs, Symptoms or Conditions	A071102	Failure to Interrogate		
							F23	Unexpected Medical Intervention	A071102	Failure to Interrogate		
					CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A071102	Failure to Interrogate		
							F23	Unexpected Medical Intervention	A071102	Failure to Interrogate		
					CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A071102	Failure to Interrogate		
							F23	Unexpected Medical Intervention	A071102	Failure to Interrogate		
					CT233	Surgical	E2403	No Clinical Signs, Symptoms or Conditions	A071102	Failure to Interrogate		
							F23	Unexpected Medical Intervention	A071102	Failure to Interrogate		
					CT2406	Implantable	E2403	No Clinical Signs, Symptoms or Conditions	A071102	Failure to Interrogate		
							F23	Unexpected Medical Intervention	A071102	Failure to Interrogate		
					CT2407	Active implantable	E2403	No Clinical Signs, Symptoms or Conditions	A071102	Failure to Interrogate		
							F23	Unexpected Medical Intervention	A071102	Failure to Interrogate		
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A071102	Failure to Interrogate		
							F23	Unexpected Medical Intervention	A071102	Failure to Interrogate		
					CT2727	Active implant devices/ancillaries	E2403	No Clinical Signs, Symptoms or Conditions	A071102	Failure to Interrogate		
							F23	Unexpected Medical Intervention	A071102	Failure to Interrogate		
					CT2750	Implantable cardiac electrical stimulators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A071102	Failure to Interrogate		
							F23	Unexpected Medical Intervention	A071102	Failure to Interrogate		
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A071102	Failure to Interrogate		
							F23	Unexpected Medical Intervention	A071102	Failure to Interrogate		
					CT2985	Active	E2403	No Clinical Signs, Symptoms or Conditions	A071102	Failure to Interrogate		
							F23	Unexpected Medical Intervention	A071102	Failure to Interrogate		
					CT3001	Implantable metal	E2403	No Clinical Signs, Symptoms or Conditions	A071102	Failure to Interrogate		
							F23	Unexpected Medical Intervention	A071102	Failure to Interrogate		
					CT3002	Implantable synthetic polymer	E2403	No Clinical Signs, Symptoms or Conditions	A071102	Failure to Interrogate		
							F23	Unexpected Medical Intervention	A071102	Failure to Interrogate		
					CT3006	Implantable inorganic material	E2403	No Clinical Signs, Symptoms or Conditions	A071102	Failure to Interrogate		
							F23	Unexpected Medical Intervention	A071102	Failure to Interrogate		
					CT314	Electrically powered	E2403	No Clinical Signs, Symptoms or Conditions	A071102	Failure to Interrogate		
							F23	Unexpected Medical Intervention	A071102	Failure to Interrogate		
					CT321	Long-term surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A071102	Failure to Interrogate		
							F23	Unexpected Medical Intervention	A071102	Failure to Interrogate		
					CT449	Generators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A071102	Failure to Interrogate		
							F23	Unexpected Medical Intervention	A071102	Failure to Interrogate		
					CT981	Single use	E2403	No Clinical Signs, Symptoms or Conditions	A071102	Failure to Interrogate		
							F23	Unexpected Medical Intervention	A071102	Failure to Interrogate		
					CT983	Surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A071102	Failure to Interrogate		
							F23	Unexpected Medical Intervention	A071102	Failure to Interrogate		
10030	2020	All other reportable incidents	47265	Dual-chamber implantable pacemaker, rate-responsive	CT141	Electrodes/Leads and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A051201	Device Dislodged or Dislocated		
							F26	No Health Consequences or Impact	A051201	Device Dislodged or Dislocated		
					CT145	Cardiovascular devices	E2403	No Clinical Signs, Symptoms or Conditions	A051201	Device Dislodged or Dislocated		
							F26	No Health Consequences or Impact	A051201	Device Dislodged or Dislocated		
					CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A051201	Device Dislodged or Dislocated		
							F26	No Health Consequences or Impact	A051201	Device Dislodged or Dislocated		
					CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A051201	Device Dislodged or Dislocated		
							F26	No Health Consequences or Impact	A051201	Device Dislodged or Dislocated		
					CT233	Surgical	E2403	No Clinical Signs, Symptoms or Conditions	A051201	Device Dislodged or Dislocated		
							F26	No Health Consequences or Impact	A051201	Device Dislodged or Dislocated		
					CT2406	Implantable	E2403	No Clinical Signs, Symptoms or Conditions	A051201	Device Dislodged or Dislocated		
							F26	No Health Consequences or Impact	A051201	Device Dislodged or Dislocated		
					CT2407	Active implantable	E2403	No Clinical Signs, Symptoms or Conditions	A051201	Device Dislodged or Dislocated		
							F26	No Health Consequences or Impact	A051201	Device Dislodged or Dislocated		
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A051201	Device Dislodged or Dislocated		
							F26	No Health Consequences or Impact	A051201	Device Dislodged or Dislocated		
					CT2727	Active implant devices/ancillaries	E2403	No Clinical Signs, Symptoms or Conditions	A051201	Device Dislodged or Dislocated		
							F26	No Health Consequences or Impact	A051201	Device Dislodged or Dislocated		
					CT2750	Implantable cardiac electrical stimulators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A051201	Device Dislodged or Dislocated		
							F26	No Health Consequences or Impact	A051201	Device Dislodged or Dislocated		
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A051201	Device Dislodged or Dislocated		
							F26	No Health Consequences or Impact	A051201	Device Dislodged or Dislocated		
					CT2985	Active	E2403	No Clinical Signs, Symptoms or Conditions	A051201	Device Dislodged or Dislocated		
							F26	No Health Consequences or Impact	A051201	Device Dislodged or Dislocated		
					CT3001	Implantable metal	E2403	No Clinical Signs, Symptoms or Conditions	A051201	Device Dislodged or Dislocated		
							F26	No Health Consequences or Impact	A051201	Device Dislodged or Dislocated		
					CT3002	Implantable synthetic polymer	E2403	No Clinical Signs, Symptoms or Conditions	A051201	Device Dislodged or Dislocated		
							F26	No Health Consequences or Impact	A051201	Device Dislodged or Dislocated		
					CT3006	Implantable inorganic material	E2403	No Clinical Signs, Symptoms or Conditions	A051201	Device Dislodged or Dislocated		
							F26	No Health Consequences or Impact	A051201	Device Dislodged or Dislocated		
					CT314	Electrically powered	E2403	No Clinical Signs, Symptoms or Conditions	A051201	Device Dislodged or Dislocated		
							F26	No Health Consequences or Impact	A051201	Device Dislodged or Dislocated		
					CT321	Long-term surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A051201	Device Dislodged or Dislocated		
							F26	No Health Consequences or Impact	A051201	Device Dislodged or Dislocated		
					CT449	Generators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A051201	Device Dislodged or Dislocated		
							F26	No Health Consequences or Impact	A051201	Device Dislodged or Dislocated		
					CT981	Single use	E2403	No Clinical Signs, Symptoms or Conditions	A051201	Device Dislodged or Dislocated		
							F26	No Health Consequences or Impact	A051201	Device Dislodged or Dislocated		
					CT983	Surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A051201	Device Dislodged or Dislocated		
							F26	No Health Consequences or Impact	A051201	Device Dislodged or Dislocated		
10031	2020	All other reportable incidents	37265	Dual-chamber implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A072201	High Impedance		
							F26	No Health Consequences or Impact	A072201	High Impedance		
10032	2020	All other reportable incidents	36241	Pacemaker/defibrillator lead	CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A040101	Fracture		
									A072201	High Impedance		
							F19	Surgical Intervention	A040101	Fracture		
									A072201	High Impedance		
10033	2019	All other reportable incidents	47265	Dual-chamber implantable pacemaker, rate-	CT141	Electrodes/Leads and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		

				responsive	CT145	Cardiovascular devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
					CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
					CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
					CT233	Surgical	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
					CT2406	Implantable	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
					CT2407	Active implantable	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
					CT2727	Active implant devices/ancillaries	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
					CT2750	Implantable cardiac electrical stimulators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
					CT2985	Active	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
					CT3001	Implantable metal	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
					CT3002	Implantable synthetic polymer	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
					CT3006	Implantable inorganic material	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
					CT314	Electrically powered	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
					CT321	Long-term surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
					CT449	Generators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
					CT981	Single use	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
					CT983	Surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
10034	2019	All other reportable incidents	47265	Dual-chamber implantable pacemaker, rate-responsive	CT141	Electrodes/Leads and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	12/11/2019	12/11/2019
					CT145	Cardiovascular devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	12/11/2019	12/11/2019
					CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	12/11/2019	12/11/2019
					CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	12/11/2019	12/11/2019
					CT233	Surgical	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	12/11/2019	12/11/2019
					CT2406	Implantable	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	12/11/2019	12/11/2019
					CT2407	Active implantable	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	12/11/2019	12/11/2019
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	12/11/2019	12/11/2019
					CT2727	Active implant devices/ancillaries	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	12/11/2019	12/11/2019
					CT2750	Implantable cardiac electrical stimulators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	12/11/2019	12/11/2019
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	12/11/2019	12/11/2019
					CT2985	Active	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	12/11/2019	12/11/2019
					CT3001	Implantable metal	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	12/11/2019	12/11/2019
					CT3002	Implantable synthetic polymer	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	12/11/2019	12/11/2019
					CT3006	Implantable inorganic material	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	12/11/2019	12/11/2019
					CT314	Electrically powered	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	12/11/2019	12/11/2019
					CT321	Long-term surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	12/11/2019	12/11/2019
					CT449	Generators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	12/11/2019	12/11/2019
					CT981	Single use	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	12/11/2019	12/11/2019
					CT983	Surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	12/11/2019	12/11/2019
10035	2019	All other reportable incidents	35852	Single-chamber implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A070909	Oversensing		
									A071302	Inappropriate/Inadequate Shock/Stimulation		
10036	2019	All other reportable incidents	36241	Pacemaker/defibrillator lead	CT1684	Cardiac pacemakers and associated devices	F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A070909	Oversensing		
									A071302	Inappropriate/Inadequate Shock/Stimulation		
10037	2019	All other reportable incidents	47267	Single-chamber implantable pacemaker, rate-responsive	CT141	Electrodes/Leads and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
					CT145	Cardiovascular devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
					CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
					CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
					CT233	Surgical	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
					CT2406	Implantable	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
					CT2407	Active implantable	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
					CT2727	Active implant devices/ancillaries	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
					CT2750	Implantable cardiac electrical stimulators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
					CT2985	Active	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
					CT3001	Implantable metal	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
					CT3002	Implantable synthetic polymer	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
					CT3006	Implantable inorganic material	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
					CT314	Electrically powered	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
					CT321	Long-term surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
					CT449	Generators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
					CT981	Single use	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
					CT983	Surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
10038	2019	All other reportable incidents	37265	Dual-chamber implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery		
10039	2019	All other reportable incidents	35223	Endocardial/interventricular septal pacing lead	CT1684	Cardiac pacemakers and associated devices	F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A070101	Failure to Capture		
10040	2019	All other reportable incidents	47265	Dual-chamber implantable pacemaker, rate-responsive	CT141	Electrodes/Leads and associated devices	F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	30/09/2019	30/09/2019
					CT145	Cardiovascular devices	F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	30/09/2019	30/09/2019
					CT1684	Cardiac pacemakers and associated devices	F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	30/09/2019	30/09/2019
					CT1685	Cardiac pacemakers	F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	30/09/2019	30/09/2019
					CT233	Surgical	F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	30/09/2019	30/09/2019
					CT2406	Implantable	F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	30/09/2019	30/09/2019
					CT2407	Active implantable	F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	30/09/2019	30/09/2019
					CT257	Implantable cardiac pacemakers	F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	30/09/2019	30/09/2019
					CT2727	Active implant devices/ancillaries	F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	30/09/2019	30/09/2019
					CT2750	Implantable cardiac electrical stimulators and associated devices	F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	30/09/2019	30/09/2019
					CT2866	Implantable cardiac electrical stimulators	F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	30/09/2019	30/09/2019
					CT2985	Active	F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	30/09/2019	30/09/2019
					CT3001	Implantable metal	F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	30/09/2019	30/09/2019
					CT3002	Implantable synthetic polymer	F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	30/09/2019	30/09/2019
					CT3006	Implantable inorganic material	F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	30/09/2019	30/09/2019
					CT314	Electrically powered	F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	30/09/2019	30/09/2019
					CT321	Long-term surgical invasive	F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	30/09/2019	30/09/2019
					CT449	Generators and associated devices	F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	30/09/2019	30/09/2019
					CT981	Single use	F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	30/09/2019	30/09/2019

					CT983	Surgical invasive	F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	30/09/2019	30/09/2019
10041	2019	All other reportable incidents	47267	Single-chamber implantable pacemaker, rate-responsive	CT141	Electrodes/Leads and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT145	Cardiovascular devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT233	Surgical	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT2406	Implantable	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT2407	Active implantable	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT2727	Active implant devices/ancillaries	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT2750	Implantable cardiac electrical stimulators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT2985	Active	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT3001	Implantable metal	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT3002	Implantable synthetic polymer	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT3006	Implantable inorganic material	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT314	Electrically powered	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT321	Long-term surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT449	Generators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT981	Single use	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT983	Surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
10042	2020	Death	37265	Dual-chamber implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E0601	Arrhythmia	A071102	Failure to Interrogate	09/09/2020	09/09/2020
									A160503	Inappropriate or Unexpected Reset	09/09/2020	09/09/2020
							E2403	No Clinical Signs, Symptoms or Conditions	A071102	Failure to Interrogate	09/09/2020	09/09/2020
									A160503	Inappropriate or Unexpected Reset	09/09/2020	09/09/2020
							F02	Death	A071102	Failure to Interrogate	09/09/2020	09/09/2020
									A160503	Inappropriate or Unexpected Reset	09/09/2020	09/09/2020
10043	2020	All other reportable incidents	35223	Endocardial/interventricular septal pacing lead	CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A0401	Break		
									A07	Electrical /Electronic Property Problem		
									A070909	Oversensing		
									A072202	Low Impedance		
							F1905	Device Revision or Replacement	A0401	Break		
									A07	Electrical /Electronic Property Problem		
									A070909	Oversensing		
									A072202	Low Impedance		
							F26	No Health Consequences or Impact	A0401	Break		
									A07	Electrical /Electronic Property Problem		
									A070909	Oversensing		
									A072202	Low Impedance		
10044	2019	All other reportable incidents	47267	Single-chamber implantable pacemaker, rate-responsive	CT141	Electrodes/Leads and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	04/12/2019	04/12/2019
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery	04/12/2019	04/12/2019
					CT145	Cardiovascular devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	04/12/2019	04/12/2019
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery	04/12/2019	04/12/2019
					CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	04/12/2019	04/12/2019
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery	04/12/2019	04/12/2019
					CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	04/12/2019	04/12/2019
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery	04/12/2019	04/12/2019
					CT233	Surgical	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	04/12/2019	04/12/2019

						F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery	04/12/2019	04/12/2019
				CT2406	Implantable	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	04/12/2019	04/12/2019
						F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery	04/12/2019	04/12/2019
				CT2407	Active implantable	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	04/12/2019	04/12/2019
						F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery	04/12/2019	04/12/2019
				CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	04/12/2019	04/12/2019
						F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery	04/12/2019	04/12/2019
				CT2727	Active implant devices/ancillaries	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	04/12/2019	04/12/2019
						F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery	04/12/2019	04/12/2019
				CT2750	Implantable cardiac electrical stimulators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	04/12/2019	04/12/2019
						F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery	04/12/2019	04/12/2019
				CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	04/12/2019	04/12/2019
						F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery	04/12/2019	04/12/2019
				CT2985	Active	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	04/12/2019	04/12/2019
						F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery	04/12/2019	04/12/2019
				CT3001	Implantable metal	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	04/12/2019	04/12/2019
						F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery	04/12/2019	04/12/2019
				CT3002	Implantable synthetic polymer	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	04/12/2019	04/12/2019
						F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery	04/12/2019	04/12/2019
				CT3006	Implantable inorganic material	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	04/12/2019	04/12/2019
						F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery	04/12/2019	04/12/2019
				CT314	Electrically powered	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	04/12/2019	04/12/2019
						F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery	04/12/2019	04/12/2019
				CT321	Long-term surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	04/12/2019	04/12/2019
						F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery	04/12/2019	04/12/2019
				CT449	Generators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	04/12/2019	04/12/2019
						F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery	04/12/2019	04/12/2019
				CT981	Single use	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	04/12/2019	04/12/2019
						F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery	04/12/2019	04/12/2019
				CT983	Surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	04/12/2019	04/12/2019
						F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery	04/12/2019	04/12/2019
10045	2020	All other reportable incidents	47265	Dual-chamber implantable pacemaker, rate-responsive	CT141	Electrodes/Leads and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A090804	False Positive Result		
							F26	No Health Consequences or Impact	A090804	False Positive Result		
					CT145	Cardiovascular devices	E2403	No Clinical Signs, Symptoms or Conditions	A090804	False Positive Result		
							F26	No Health Consequences or Impact	A090804	False Positive Result		
					CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A090804	False Positive Result		
							F26	No Health Consequences or Impact	A090804	False Positive Result		
					CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A090804	False Positive Result		
							F26	No Health Consequences or Impact	A090804	False Positive Result		
					CT233	Surgical	E2403	No Clinical Signs, Symptoms or Conditions	A090804	False Positive Result		
							F26	No Health Consequences or Impact	A090804	False Positive Result		
					CT2406	Implantable	E2403	No Clinical Signs, Symptoms or Conditions	A090804	False Positive Result		
							F26	No Health Consequences or Impact	A090804	False Positive Result		
					CT2407	Active implantable	E2403	No Clinical Signs, Symptoms or Conditions	A090804	False Positive Result		
							F26	No Health Consequences or Impact	A090804	False Positive Result		
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A090804	False Positive Result		
							F26	No Health Consequences or Impact	A090804	False Positive Result		
					CT2727	Active implant devices/ancillaries	E2403	No Clinical Signs, Symptoms or Conditions	A090804	False Positive Result		
							F26	No Health Consequences or Impact	A090804	False Positive Result		
					CT2750	Implantable cardiac electrical stimulators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A090804	False Positive Result		
							F26	No Health Consequences or Impact	A090804	False Positive Result		
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A090804	False Positive Result		
							F26	No Health Consequences or Impact	A090804	False Positive Result		
					CT2985	Active	E2403	No Clinical Signs, Symptoms or Conditions	A090804	False Positive Result		
							F26	No Health Consequences or Impact	A090804	False Positive Result		
					CT3001	Implantable metal	E2403	No Clinical Signs, Symptoms or Conditions	A090804	False Positive Result		
							F26	No Health Consequences or Impact	A090804	False Positive Result		
					CT3002	Implantable synthetic polymer	E2403	No Clinical Signs, Symptoms or Conditions	A090804	False Positive Result		
							F26	No Health Consequences or Impact	A090804	False Positive Result		
					CT3006	Implantable inorganic material	E2403	No Clinical Signs, Symptoms or Conditions	A090804	False Positive Result		
							F26	No Health Consequences or Impact	A090804	False Positive Result		
					CT314	Electrically powered	E2403	No Clinical Signs, Symptoms or Conditions	A090804	False Positive Result		
							F26	No Health Consequences or Impact	A090804	False Positive Result		
					CT321	Long-term surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A090804	False Positive Result		
							F26	No Health Consequences or Impact	A090804	False Positive Result		
					CT449	Generators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A090804	False Positive Result		
							F26	No Health Consequences or Impact	A090804	False Positive Result		
					CT981	Single use	E2403	No Clinical Signs, Symptoms or Conditions	A090804	False Positive Result		
							F26	No Health Consequences or Impact	A090804	False Positive Result		
					CT983	Surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A090804	False Positive Result		
							F26	No Health Consequences or Impact	A090804	False Positive Result		
10046	2019	All other reportable incidents	35852	Single-chamber implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A070909	Oversensing		
10047	2020	All other reportable incidents	35223	Endocardial/interventricular septal pacing lead	CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A0722	Impedance Problem		
									A072201	High Impedance		
							F26	No Health Consequences or Impact	A0722	Impedance Problem		
									A072201	High Impedance		
10048	2020	All other reportable incidents	47265	Dual-chamber implantable pacemaker, rate-responsive	CT141	Electrodes/Leads and associated devices	E060102	Atrial Fibrillation	A0712	Pacing Problem		
									A071202	Inaccurate Synchronization		
									A090801	Signal Artifact/Noise		
							E2403	No Clinical Signs, Symptoms or Conditions	A0712	Pacing Problem		
									A071202	Inaccurate Synchronization		
									A090801	Signal Artifact/Noise		
							F26	No Health Consequences or Impact	A0712	Pacing Problem		
									A071202	Inaccurate Synchronization		
									A090801	Signal Artifact/Noise		
					CT145	Cardiovascular devices	E060102	Atrial Fibrillation	A0712	Pacing Problem		
									A071202	Inaccurate Synchronization		
									A090801	Signal Artifact/Noise		

						E2403	No Clinical Signs, Symptoms or Conditions	A0712	Pacing Problem		
								A071202	Inaccurate Synchronization		
								A090801	Signal Artifact/Noise		
						F26	No Health Consequences or Impact	A0712	Pacing Problem		
								A071202	Inaccurate Synchronization		
								A090801	Signal Artifact/Noise		
					CT1684	Cardiac pacemakers and associated devices	E060102	Atrial Fibrillation	A0712	Pacing Problem		
									A071202	Inaccurate Synchronization		
									A090801	Signal Artifact/Noise		
							E2403	No Clinical Signs, Symptoms or Conditions	A0712	Pacing Problem		
									A071202	Inaccurate Synchronization		
									A090801	Signal Artifact/Noise		
							F26	No Health Consequences or Impact	A0712	Pacing Problem		
									A071202	Inaccurate Synchronization		
									A090801	Signal Artifact/Noise		
					CT1685	Cardiac pacemakers	E060102	Atrial Fibrillation	A0712	Pacing Problem		
									A071202	Inaccurate Synchronization		
									A090801	Signal Artifact/Noise		
							E2403	No Clinical Signs, Symptoms or Conditions	A0712	Pacing Problem		
									A071202	Inaccurate Synchronization		
									A090801	Signal Artifact/Noise		
							F26	No Health Consequences or Impact	A0712	Pacing Problem		
									A071202	Inaccurate Synchronization		
									A090801	Signal Artifact/Noise		
					CT233	Surgical	E060102	Atrial Fibrillation	A0712	Pacing Problem		
									A071202	Inaccurate Synchronization		
									A090801	Signal Artifact/Noise		
							E2403	No Clinical Signs, Symptoms or Conditions	A0712	Pacing Problem		
									A071202	Inaccurate Synchronization		
									A090801	Signal Artifact/Noise		
							F26	No Health Consequences or Impact	A0712	Pacing Problem		
									A071202	Inaccurate Synchronization		
									A090801	Signal Artifact/Noise		
					CT2406	Implantable	E060102	Atrial Fibrillation	A0712	Pacing Problem		
									A071202	Inaccurate Synchronization		
									A090801	Signal Artifact/Noise		
							E2403	No Clinical Signs, Symptoms or Conditions	A0712	Pacing Problem		
									A071202	Inaccurate Synchronization		
									A090801	Signal Artifact/Noise		
							F26	No Health Consequences or Impact	A0712	Pacing Problem		
									A071202	Inaccurate Synchronization		
									A090801	Signal Artifact/Noise		
					CT2407	Active implantable	E060102	Atrial Fibrillation	A0712	Pacing Problem		
									A071202	Inaccurate Synchronization		
									A090801	Signal Artifact/Noise		
							E2403	No Clinical Signs, Symptoms or Conditions	A0712	Pacing Problem		
									A071202	Inaccurate Synchronization		
									A090801	Signal Artifact/Noise		
							F26	No Health Consequences or Impact	A0712	Pacing Problem		
									A071202	Inaccurate Synchronization		
									A090801	Signal Artifact/Noise		
					CT257	Implantable cardiac pacemakers	E060102	Atrial Fibrillation	A0712	Pacing Problem		
									A071202	Inaccurate Synchronization		
									A090801	Signal Artifact/Noise		
							E2403	No Clinical Signs, Symptoms or Conditions	A0712	Pacing Problem		
									A071202	Inaccurate Synchronization		
									A090801	Signal Artifact/Noise		
							F26	No Health Consequences or Impact	A0712	Pacing Problem		
									A071202	Inaccurate Synchronization		
									A090801	Signal Artifact/Noise		
					CT2727	Active implant devices/ancillaries	E060102	Atrial Fibrillation	A0712	Pacing Problem		
									A071202	Inaccurate Synchronization		
									A090801	Signal Artifact/Noise		
							E2403	No Clinical Signs, Symptoms or Conditions	A0712	Pacing Problem		
									A071202	Inaccurate Synchronization		
									A090801	Signal Artifact/Noise		
							F26	No Health Consequences or Impact	A0712	Pacing Problem		
									A071202	Inaccurate Synchronization		
									A090801	Signal Artifact/Noise		
					CT2750	Implantable cardiac electrical stimulators and associated devices	E060102	Atrial Fibrillation	A0712	Pacing Problem		
									A071202	Inaccurate Synchronization		
									A090801	Signal Artifact/Noise		
							E2403	No Clinical Signs, Symptoms or Conditions	A0712	Pacing Problem		
									A071202	Inaccurate Synchronization		
									A090801	Signal Artifact/Noise		
							F26	No Health Consequences or Impact	A0712	Pacing Problem		
									A071202	Inaccurate Synchronization		
									A090801	Signal Artifact/Noise		
					CT2866	Implantable cardiac electrical stimulators	E060102	Atrial Fibrillation	A0712	Pacing Problem		
									A071202	Inaccurate Synchronization		
									A090801	Signal Artifact/Noise		
							E2403	No Clinical Signs, Symptoms or Conditions	A0712	Pacing Problem		
									A071202	Inaccurate Synchronization		
									A090801	Signal Artifact/Noise		
							F26	No Health Consequences or Impact	A0712	Pacing Problem		
									A071202	Inaccurate Synchronization		
									A090801	Signal Artifact/Noise		
					CT2985	Active	E060102	Atrial Fibrillation	A0712	Pacing Problem		

									A071202	Inaccurate Synchronization		
									A090801	Signal Artifact/Noise		
							E2403	No Clinical Signs, Symptoms or Conditions	A0712	Pacing Problem		
									A071202	Inaccurate Synchronization		
									A090801	Signal Artifact/Noise		
							F26	No Health Consequences or Impact	A0712	Pacing Problem		
									A071202	Inaccurate Synchronization		
									A090801	Signal Artifact/Noise		
					CT3001	Implantable metal	E060102	Atrial Fibrillation	A0712	Pacing Problem		
									A071202	Inaccurate Synchronization		
									A090801	Signal Artifact/Noise		
							E2403	No Clinical Signs, Symptoms or Conditions	A0712	Pacing Problem		
									A071202	Inaccurate Synchronization		
									A090801	Signal Artifact/Noise		
							F26	No Health Consequences or Impact	A0712	Pacing Problem		
									A071202	Inaccurate Synchronization		
									A090801	Signal Artifact/Noise		
					CT3002	Implantable synthetic polymer	E060102	Atrial Fibrillation	A0712	Pacing Problem		
									A071202	Inaccurate Synchronization		
									A090801	Signal Artifact/Noise		
							E2403	No Clinical Signs, Symptoms or Conditions	A0712	Pacing Problem		
									A071202	Inaccurate Synchronization		
									A090801	Signal Artifact/Noise		
							F26	No Health Consequences or Impact	A0712	Pacing Problem		
									A071202	Inaccurate Synchronization		
									A090801	Signal Artifact/Noise		
					CT3006	Implantable inorganic material	E060102	Atrial Fibrillation	A0712	Pacing Problem		
									A071202	Inaccurate Synchronization		
									A090801	Signal Artifact/Noise		
							E2403	No Clinical Signs, Symptoms or Conditions	A0712	Pacing Problem		
									A071202	Inaccurate Synchronization		
									A090801	Signal Artifact/Noise		
							F26	No Health Consequences or Impact	A0712	Pacing Problem		
									A071202	Inaccurate Synchronization		
									A090801	Signal Artifact/Noise		
					CT314	Electrically powered	E060102	Atrial Fibrillation	A0712	Pacing Problem		
									A071202	Inaccurate Synchronization		
									A090801	Signal Artifact/Noise		
							E2403	No Clinical Signs, Symptoms or Conditions	A0712	Pacing Problem		
									A071202	Inaccurate Synchronization		
									A090801	Signal Artifact/Noise		
							F26	No Health Consequences or Impact	A0712	Pacing Problem		
									A071202	Inaccurate Synchronization		
									A090801	Signal Artifact/Noise		
					CT321	Long-term surgical invasive	E060102	Atrial Fibrillation	A0712	Pacing Problem		
									A071202	Inaccurate Synchronization		
									A090801	Signal Artifact/Noise		
							E2403	No Clinical Signs, Symptoms or Conditions	A0712	Pacing Problem		
									A071202	Inaccurate Synchronization		
									A090801	Signal Artifact/Noise		
							F26	No Health Consequences or Impact	A0712	Pacing Problem		
									A071202	Inaccurate Synchronization		
									A090801	Signal Artifact/Noise		
					CT449	Generators and associated devices	E060102	Atrial Fibrillation	A0712	Pacing Problem		
									A071202	Inaccurate Synchronization		
									A090801	Signal Artifact/Noise		
							E2403	No Clinical Signs, Symptoms or Conditions	A0712	Pacing Problem		
									A071202	Inaccurate Synchronization		
									A090801	Signal Artifact/Noise		
							F26	No Health Consequences or Impact	A0712	Pacing Problem		
									A071202	Inaccurate Synchronization		
									A090801	Signal Artifact/Noise		
					CT981	Single use	E060102	Atrial Fibrillation	A0712	Pacing Problem		
									A071202	Inaccurate Synchronization		
									A090801	Signal Artifact/Noise		
							E2403	No Clinical Signs, Symptoms or Conditions	A0712	Pacing Problem		
									A071202	Inaccurate Synchronization		
									A090801	Signal Artifact/Noise		
							F26	No Health Consequences or Impact	A0712	Pacing Problem		
									A071202	Inaccurate Synchronization		
									A090801	Signal Artifact/Noise		
					CT983	Surgical invasive	E060102	Atrial Fibrillation	A0712	Pacing Problem		
									A071202	Inaccurate Synchronization		
									A090801	Signal Artifact/Noise		
							E2403	No Clinical Signs, Symptoms or Conditions	A0712	Pacing Problem		
									A071202	Inaccurate Synchronization		
									A090801	Signal Artifact/Noise		
							F26	No Health Consequences or Impact	A0712	Pacing Problem		
									A071202	Inaccurate Synchronization		
									A090801	Signal Artifact/Noise		
10049	2019	All other reportable incidents	47265	Dual-chamber implantable pacemaker, rate-responsive	CT141	Electrodes/Leads and associated devices	F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	20/09/2019	
					CT145	Cardiovascular devices	F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	20/09/2019	
					CT1684	Cardiac pacemakers and associated devices	F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	20/09/2019	
					CT1685	Cardiac pacemakers	F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	20/09/2019	
					CT233	Surgical	F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	20/09/2019	
					CT2406	Implantable	F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	20/09/2019	
					CT2407	Active implantable	F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	20/09/2019	
					CT257	Implantable cardiac pacemakers	F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	20/09/2019	

					CT2727	Active implant devices/ancillaries	F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	20/09/2019	
					CT2750	Implantable cardiac electrical stimulators and associated devices	F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	20/09/2019	
					CT2866	Implantable cardiac electrical stimulators	F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	20/09/2019	
					CT2985	Active	F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	20/09/2019	
					CT3001	Implantable metal	F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	20/09/2019	
					CT3002	Implantable synthetic polymer	F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	20/09/2019	
					CT3006	Implantable inorganic material	F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	20/09/2019	
					CT314	Electrically powered	F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	20/09/2019	
					CT321	Long-term surgical invasive	F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	20/09/2019	
					CT449	Generators and associated devices	F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	20/09/2019	
					CT981	Single use	F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	20/09/2019	
					CT983	Surgical invasive	F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	20/09/2019	
10050	2019	All other reportable incidents	47270	Cardiac resynchronization therapy implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	F1905	Device Revision or Replacement	A0705	Battery Problem		
10051	2020	All other reportable incidents	47265	Dual-chamber implantable pacemaker, rate-responsive	CT141	Electrodes/Leads and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem	13/10/2020	13/10/2020
									A070504	Premature Discharge of Battery	13/10/2020	13/10/2020
							F08	Hospitalization or Prolonged Hospitalization	A0705	Battery Problem	13/10/2020	13/10/2020
									A070504	Premature Discharge of Battery	13/10/2020	13/10/2020
							F1903	Device Explantation	A0705	Battery Problem	13/10/2020	13/10/2020
									A070504	Premature Discharge of Battery	13/10/2020	13/10/2020
					CT145	Cardiovascular devices	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem	13/10/2020	13/10/2020
									A070504	Premature Discharge of Battery	13/10/2020	13/10/2020
							F08	Hospitalization or Prolonged Hospitalization	A0705	Battery Problem	13/10/2020	13/10/2020
									A070504	Premature Discharge of Battery	13/10/2020	13/10/2020
							F1903	Device Explantation	A0705	Battery Problem	13/10/2020	13/10/2020
									A070504	Premature Discharge of Battery	13/10/2020	13/10/2020
					CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem	13/10/2020	13/10/2020
									A070504	Premature Discharge of Battery	13/10/2020	13/10/2020
							F08	Hospitalization or Prolonged Hospitalization	A0705	Battery Problem	13/10/2020	13/10/2020
									A070504	Premature Discharge of Battery	13/10/2020	13/10/2020
							F1903	Device Explantation	A0705	Battery Problem	13/10/2020	13/10/2020
									A070504	Premature Discharge of Battery	13/10/2020	13/10/2020
					CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem	13/10/2020	13/10/2020
									A070504	Premature Discharge of Battery	13/10/2020	13/10/2020
							F08	Hospitalization or Prolonged Hospitalization	A0705	Battery Problem	13/10/2020	13/10/2020
									A070504	Premature Discharge of Battery	13/10/2020	13/10/2020
							F1903	Device Explantation	A0705	Battery Problem	13/10/2020	13/10/2020
									A070504	Premature Discharge of Battery	13/10/2020	13/10/2020
					CT233	Surgical	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem	13/10/2020	13/10/2020
									A070504	Premature Discharge of Battery	13/10/2020	13/10/2020
							F08	Hospitalization or Prolonged Hospitalization	A0705	Battery Problem	13/10/2020	13/10/2020
									A070504	Premature Discharge of Battery	13/10/2020	13/10/2020
							F1903	Device Explantation	A0705	Battery Problem	13/10/2020	13/10/2020
									A070504	Premature Discharge of Battery	13/10/2020	13/10/2020
					CT2406	Implantable	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem	13/10/2020	13/10/2020
									A070504	Premature Discharge of Battery	13/10/2020	13/10/2020
							F08	Hospitalization or Prolonged Hospitalization	A0705	Battery Problem	13/10/2020	13/10/2020
									A070504	Premature Discharge of Battery	13/10/2020	13/10/2020
							F1903	Device Explantation	A0705	Battery Problem	13/10/2020	13/10/2020
									A070504	Premature Discharge of Battery	13/10/2020	13/10/2020
					CT2407	Active implantable	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem	13/10/2020	13/10/2020
									A070504	Premature Discharge of Battery	13/10/2020	13/10/2020
							F08	Hospitalization or Prolonged Hospitalization	A0705	Battery Problem	13/10/2020	13/10/2020
									A070504	Premature Discharge of Battery	13/10/2020	13/10/2020
							F1903	Device Explantation	A0705	Battery Problem	13/10/2020	13/10/2020
									A070504	Premature Discharge of Battery	13/10/2020	13/10/2020
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem	13/10/2020	13/10/2020
									A070504	Premature Discharge of Battery	13/10/2020	13/10/2020
							F08	Hospitalization or Prolonged Hospitalization	A0705	Battery Problem	13/10/2020	13/10/2020
									A070504	Premature Discharge of Battery	13/10/2020	13/10/2020
							F1903	Device Explantation	A0705	Battery Problem	13/10/2020	13/10/2020
									A070504	Premature Discharge of Battery	13/10/2020	13/10/2020
					CT2727	Active implant devices/ancillaries	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem	13/10/2020	13/10/2020
									A070504	Premature Discharge of Battery	13/10/2020	13/10/2020
							F08	Hospitalization or Prolonged Hospitalization	A0705	Battery Problem	13/10/2020	13/10/2020
									A070504	Premature Discharge of Battery	13/10/2020	13/10/2020
							F1903	Device Explantation	A0705	Battery Problem	13/10/2020	13/10/2020
									A070504	Premature Discharge of Battery	13/10/2020	13/10/2020
					CT2750	Implantable cardiac electrical stimulators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem	13/10/2020	13/10/2020
									A070504	Premature Discharge of Battery	13/10/2020	13/10/2020
							F08	Hospitalization or Prolonged Hospitalization	A0705	Battery Problem	13/10/2020	13/10/2020
									A070504	Premature Discharge of Battery	13/10/2020	13/10/2020
							F1903	Device Explantation	A0705	Battery Problem	13/10/2020	13/10/2020
									A070504	Premature Discharge of Battery	13/10/2020	13/10/2020
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem	13/10/2020	13/10/2020
									A070504	Premature Discharge of Battery	13/10/2020	13/10/2020
							F08	Hospitalization or Prolonged Hospitalization	A0705	Battery Problem	13/10/2020	13/10/2020
									A070504	Premature Discharge of Battery	13/10/2020	13/10/2020
							F1903	Device Explantation	A0705	Battery Problem	13/10/2020	13/10/2020
									A070504	Premature Discharge of Battery	13/10/2020	13/10/2020
					CT2985	Active	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem	13/10/2020	13/10/2020
									A070504	Premature Discharge of Battery	13/10/2020	13/10/2020
							F08	Hospitalization or Prolonged Hospitalization	A0705	Battery Problem	13/10/2020	13/10/2020
									A070504	Premature Discharge of Battery	13/10/2020	13/10/2020
							F1903	Device Explantation	A0705	Battery Problem	13/10/2020	13/10/2020
									A070504	Premature Discharge of Battery	13/10/2020	13/10/2020
					CT3001	Implantable metal	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem	13/10/2020	13/10/2020

									A070504	Premature Discharge of Battery	13/10/2020	13/10/2020
							F08	Hospitalization or Prolonged Hospitalization	A0705	Battery Problem	13/10/2020	13/10/2020
									A070504	Premature Discharge of Battery	13/10/2020	13/10/2020
							F1903	Device Explantation	A0705	Battery Problem	13/10/2020	13/10/2020
									A070504	Premature Discharge of Battery	13/10/2020	13/10/2020
					CT3002	Implantable synthetic polymer	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem	13/10/2020	13/10/2020
									A070504	Premature Discharge of Battery	13/10/2020	13/10/2020
							F08	Hospitalization or Prolonged Hospitalization	A0705	Battery Problem	13/10/2020	13/10/2020
									A070504	Premature Discharge of Battery	13/10/2020	13/10/2020
							F1903	Device Explantation	A0705	Battery Problem	13/10/2020	13/10/2020
									A070504	Premature Discharge of Battery	13/10/2020	13/10/2020
					CT3006	Implantable inorganic material	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem	13/10/2020	13/10/2020
									A070504	Premature Discharge of Battery	13/10/2020	13/10/2020
							F08	Hospitalization or Prolonged Hospitalization	A0705	Battery Problem	13/10/2020	13/10/2020
									A070504	Premature Discharge of Battery	13/10/2020	13/10/2020
							F1903	Device Explantation	A0705	Battery Problem	13/10/2020	13/10/2020
									A070504	Premature Discharge of Battery	13/10/2020	13/10/2020
					CT314	Electrically powered	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem	13/10/2020	13/10/2020
									A070504	Premature Discharge of Battery	13/10/2020	13/10/2020
							F08	Hospitalization or Prolonged Hospitalization	A0705	Battery Problem	13/10/2020	13/10/2020
									A070504	Premature Discharge of Battery	13/10/2020	13/10/2020
							F1903	Device Explantation	A0705	Battery Problem	13/10/2020	13/10/2020
									A070504	Premature Discharge of Battery	13/10/2020	13/10/2020
					CT321	Long-term surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem	13/10/2020	13/10/2020
									A070504	Premature Discharge of Battery	13/10/2020	13/10/2020
							F08	Hospitalization or Prolonged Hospitalization	A0705	Battery Problem	13/10/2020	13/10/2020
									A070504	Premature Discharge of Battery	13/10/2020	13/10/2020
							F1903	Device Explantation	A0705	Battery Problem	13/10/2020	13/10/2020
									A070504	Premature Discharge of Battery	13/10/2020	13/10/2020
					CT449	Generators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem	13/10/2020	13/10/2020
									A070504	Premature Discharge of Battery	13/10/2020	13/10/2020
							F08	Hospitalization or Prolonged Hospitalization	A0705	Battery Problem	13/10/2020	13/10/2020
									A070504	Premature Discharge of Battery	13/10/2020	13/10/2020
							F1903	Device Explantation	A0705	Battery Problem	13/10/2020	13/10/2020
									A070504	Premature Discharge of Battery	13/10/2020	13/10/2020
					CT981	Single use	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem	13/10/2020	13/10/2020
									A070504	Premature Discharge of Battery	13/10/2020	13/10/2020
							F08	Hospitalization or Prolonged Hospitalization	A0705	Battery Problem	13/10/2020	13/10/2020
									A070504	Premature Discharge of Battery	13/10/2020	13/10/2020
							F1903	Device Explantation	A0705	Battery Problem	13/10/2020	13/10/2020
									A070504	Premature Discharge of Battery	13/10/2020	13/10/2020
					CT983	Surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem	13/10/2020	13/10/2020
									A070504	Premature Discharge of Battery	13/10/2020	13/10/2020
							F08	Hospitalization or Prolonged Hospitalization	A0705	Battery Problem	13/10/2020	13/10/2020
									A070504	Premature Discharge of Battery	13/10/2020	13/10/2020
							F1903	Device Explantation	A0705	Battery Problem	13/10/2020	13/10/2020
									A070504	Premature Discharge of Battery	13/10/2020	13/10/2020
10052	2020	All other reportable incidents	36241	Pacemaker/defibrillator lead	CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A040101	Fracture	18/09/2020	18/09/2020
									A072201	High Impedance	18/09/2020	18/09/2020
							F08	Hospitalization or Prolonged Hospitalization	A040101	Fracture	18/09/2020	18/09/2020
									A072201	High Impedance	18/09/2020	18/09/2020
							F12	Serious Injury/ Illness/ Impairment	A040101	Fracture	18/09/2020	18/09/2020
									A072201	High Impedance	18/09/2020	18/09/2020
							F1905	Device Revision or Replacement	A040101	Fracture	18/09/2020	18/09/2020
									A072201	High Impedance	18/09/2020	18/09/2020
10053	2019	All other reportable incidents	35852	Single-chamber implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A070909	Oversensing		
10054	2019	All other reportable incidents	35223	Endocardial/interventricular septal pacing lead	CT1684	Cardiac pacemakers and associated devices	E2114	Perforation	A24	No Apparent Device Problem		
10055	2019	All other reportable incidents	37265	Dual-chamber implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A071301	Failure to Deliver Shock/Stimulation	30/10/2019	
10056	2019	All other reportable incidents	47270	Cardiac resynchronization therapy implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery		
10057	2019	All other reportable incidents	47265	Dual-chamber implantable pacemaker, rate-responsive	CT141	Electrodes/Leads and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
					CT145	Cardiovascular devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
					CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
					CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
					CT233	Surgical	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
					CT2406	Implantable	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
					CT2407	Active implantable	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
					CT2727	Active implant devices/ancillaries	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
					CT2750	Implantable cardiac electrical stimulators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
					CT2985	Active	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
					CT3001	Implantable metal	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
					CT3002	Implantable synthetic polymer	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
					CT3006	Implantable inorganic material	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
					CT314	Electrically powered	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
					CT321	Long-term surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
					CT449	Generators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
					CT981	Single use	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
					CT983	Surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
10058	2020	All other reportable incidents	47265	Dual-chamber implantable pacemaker, rate-responsive	CT141	Electrodes/Leads and associated devices	E0601	Arrhythmia	A070504	Premature Discharge of Battery	18/09/2020	18/09/2020
									A071102	Failure to Interrogate	18/09/2020	18/09/2020
									A071204	No Pacing	18/09/2020	18/09/2020
							F1901	Additional Surgery	A070504	Premature Discharge of Battery	18/09/2020	18/09/2020
									A071102	Failure to Interrogate	18/09/2020	18/09/2020
									A071204	No Pacing	18/09/2020	18/09/2020

							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	18/09/2020	18/09/2020
									A071102	Failure to Interrogate	18/09/2020	18/09/2020
									A071204	No Pacing	18/09/2020	18/09/2020
					CT145	Cardiovascular devices	E0601	Arrhythmia	A070504	Premature Discharge of Battery	18/09/2020	18/09/2020
									A071102	Failure to Interrogate	18/09/2020	18/09/2020
									A071204	No Pacing	18/09/2020	18/09/2020
							F1901	Additional Surgery	A070504	Premature Discharge of Battery	18/09/2020	18/09/2020
									A071102	Failure to Interrogate	18/09/2020	18/09/2020
									A071204	No Pacing	18/09/2020	18/09/2020
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	18/09/2020	18/09/2020
									A071102	Failure to Interrogate	18/09/2020	18/09/2020
									A071204	No Pacing	18/09/2020	18/09/2020
					CT1684	Cardiac pacemakers and associated devices	E0601	Arrhythmia	A070504	Premature Discharge of Battery	18/09/2020	18/09/2020
									A071102	Failure to Interrogate	18/09/2020	18/09/2020
									A071204	No Pacing	18/09/2020	18/09/2020
							F1901	Additional Surgery	A070504	Premature Discharge of Battery	18/09/2020	18/09/2020
									A071102	Failure to Interrogate	18/09/2020	18/09/2020
									A071204	No Pacing	18/09/2020	18/09/2020
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	18/09/2020	18/09/2020
									A071102	Failure to Interrogate	18/09/2020	18/09/2020
									A071204	No Pacing	18/09/2020	18/09/2020
					CT1685	Cardiac pacemakers	E0601	Arrhythmia	A070504	Premature Discharge of Battery	18/09/2020	18/09/2020
									A071102	Failure to Interrogate	18/09/2020	18/09/2020
									A071204	No Pacing	18/09/2020	18/09/2020
							F1901	Additional Surgery	A070504	Premature Discharge of Battery	18/09/2020	18/09/2020
									A071102	Failure to Interrogate	18/09/2020	18/09/2020
									A071204	No Pacing	18/09/2020	18/09/2020
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	18/09/2020	18/09/2020
									A071102	Failure to Interrogate	18/09/2020	18/09/2020
									A071204	No Pacing	18/09/2020	18/09/2020
					CT233	Surgical	E0601	Arrhythmia	A070504	Premature Discharge of Battery	18/09/2020	18/09/2020
									A071102	Failure to Interrogate	18/09/2020	18/09/2020
									A071204	No Pacing	18/09/2020	18/09/2020
							F1901	Additional Surgery	A070504	Premature Discharge of Battery	18/09/2020	18/09/2020
									A071102	Failure to Interrogate	18/09/2020	18/09/2020
									A071204	No Pacing	18/09/2020	18/09/2020
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	18/09/2020	18/09/2020
									A071102	Failure to Interrogate	18/09/2020	18/09/2020
									A071204	No Pacing	18/09/2020	18/09/2020
					CT2406	Implantable	E0601	Arrhythmia	A070504	Premature Discharge of Battery	18/09/2020	18/09/2020
									A071102	Failure to Interrogate	18/09/2020	18/09/2020
									A071204	No Pacing	18/09/2020	18/09/2020
							F1901	Additional Surgery	A070504	Premature Discharge of Battery	18/09/2020	18/09/2020
									A071102	Failure to Interrogate	18/09/2020	18/09/2020
									A071204	No Pacing	18/09/2020	18/09/2020
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	18/09/2020	18/09/2020
									A071102	Failure to Interrogate	18/09/2020	18/09/2020
									A071204	No Pacing	18/09/2020	18/09/2020
					CT2407	Active implantable	E0601	Arrhythmia	A070504	Premature Discharge of Battery	18/09/2020	18/09/2020
									A071102	Failure to Interrogate	18/09/2020	18/09/2020
									A071204	No Pacing	18/09/2020	18/09/2020
							F1901	Additional Surgery	A070504	Premature Discharge of Battery	18/09/2020	18/09/2020
									A071102	Failure to Interrogate	18/09/2020	18/09/2020
									A071204	No Pacing	18/09/2020	18/09/2020
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	18/09/2020	18/09/2020
									A071102	Failure to Interrogate	18/09/2020	18/09/2020
									A071204	No Pacing	18/09/2020	18/09/2020
					CT257	Implantable cardiac pacemakers	E0601	Arrhythmia	A070504	Premature Discharge of Battery	18/09/2020	18/09/2020
									A071102	Failure to Interrogate	18/09/2020	18/09/2020
									A071204	No Pacing	18/09/2020	18/09/2020
							F1901	Additional Surgery	A070504	Premature Discharge of Battery	18/09/2020	18/09/2020
									A071102	Failure to Interrogate	18/09/2020	18/09/2020
									A071204	No Pacing	18/09/2020	18/09/2020
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	18/09/2020	18/09/2020
									A071102	Failure to Interrogate	18/09/2020	18/09/2020
									A071204	No Pacing	18/09/2020	18/09/2020
					CT2727	Active implant devices/ancillaries	E0601	Arrhythmia	A070504	Premature Discharge of Battery	18/09/2020	18/09/2020
									A071102	Failure to Interrogate	18/09/2020	18/09/2020
									A071204	No Pacing	18/09/2020	18/09/2020
							F1901	Additional Surgery	A070504	Premature Discharge of Battery	18/09/2020	18/09/2020
									A071102	Failure to Interrogate	18/09/2020	18/09/2020
									A071204	No Pacing	18/09/2020	18/09/2020
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	18/09/2020	18/09/2020
									A071102	Failure to Interrogate	18/09/2020	18/09/2020
									A071204	No Pacing	18/09/2020	18/09/2020
					CT2750	Implantable cardiac electrical stimulators and associated devices	E0601	Arrhythmia	A070504	Premature Discharge of Battery	18/09/2020	18/09/2020
									A071102	Failure to Interrogate	18/09/2020	18/09/2020
									A071204	No Pacing	18/09/2020	18/09/2020
							F1901	Additional Surgery	A070504	Premature Discharge of Battery	18/09/2020	18/09/2020
									A071102	Failure to Interrogate	18/09/2020	18/09/2020
									A071204	No Pacing	18/09/2020	18/09/2020
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	18/09/2020	18/09/2020
									A071102	Failure to Interrogate	18/09/2020	18/09/2020
									A071204	No Pacing	18/09/2020	18/09/2020
					CT2866	Implantable cardiac electrical stimulators	E0601	Arrhythmia	A070504	Premature Discharge of Battery	18/09/2020	18/09/2020
									A071102	Failure to Interrogate	18/09/2020	18/09/2020
									A071204	No Pacing	18/09/2020	18/09/2020
							F1901	Additional Surgery	A070504	Premature Discharge of Battery	18/09/2020	18/09/2020

									A071102	Failure to Interrogate	18/09/2020	18/09/2020
									A071204	No Pacing	18/09/2020	18/09/2020
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	18/09/2020	18/09/2020
									A071102	Failure to Interrogate	18/09/2020	18/09/2020
									A071204	No Pacing	18/09/2020	18/09/2020
					CT2985	Active	E0601	Arrhythmia	A070504	Premature Discharge of Battery	18/09/2020	18/09/2020
									A071102	Failure to Interrogate	18/09/2020	18/09/2020
									A071204	No Pacing	18/09/2020	18/09/2020
							F1901	Additional Surgery	A070504	Premature Discharge of Battery	18/09/2020	18/09/2020
									A071102	Failure to Interrogate	18/09/2020	18/09/2020
									A071204	No Pacing	18/09/2020	18/09/2020
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	18/09/2020	18/09/2020
									A071102	Failure to Interrogate	18/09/2020	18/09/2020
									A071204	No Pacing	18/09/2020	18/09/2020
					CT3001	Implantable metal	E0601	Arrhythmia	A070504	Premature Discharge of Battery	18/09/2020	18/09/2020
									A071102	Failure to Interrogate	18/09/2020	18/09/2020
									A071204	No Pacing	18/09/2020	18/09/2020
							F1901	Additional Surgery	A070504	Premature Discharge of Battery	18/09/2020	18/09/2020
									A071102	Failure to Interrogate	18/09/2020	18/09/2020
									A071204	No Pacing	18/09/2020	18/09/2020
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	18/09/2020	18/09/2020
									A071102	Failure to Interrogate	18/09/2020	18/09/2020
									A071204	No Pacing	18/09/2020	18/09/2020
					CT3002	Implantable synthetic polymer	E0601	Arrhythmia	A070504	Premature Discharge of Battery	18/09/2020	18/09/2020
									A071102	Failure to Interrogate	18/09/2020	18/09/2020
									A071204	No Pacing	18/09/2020	18/09/2020
							F1901	Additional Surgery	A070504	Premature Discharge of Battery	18/09/2020	18/09/2020
									A071102	Failure to Interrogate	18/09/2020	18/09/2020
									A071204	No Pacing	18/09/2020	18/09/2020
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	18/09/2020	18/09/2020
									A071102	Failure to Interrogate	18/09/2020	18/09/2020
									A071204	No Pacing	18/09/2020	18/09/2020
					CT3006	Implantable inorganic material	E0601	Arrhythmia	A070504	Premature Discharge of Battery	18/09/2020	18/09/2020
									A071102	Failure to Interrogate	18/09/2020	18/09/2020
									A071204	No Pacing	18/09/2020	18/09/2020
							F1901	Additional Surgery	A070504	Premature Discharge of Battery	18/09/2020	18/09/2020
									A071102	Failure to Interrogate	18/09/2020	18/09/2020
									A071204	No Pacing	18/09/2020	18/09/2020
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	18/09/2020	18/09/2020
									A071102	Failure to Interrogate	18/09/2020	18/09/2020
									A071204	No Pacing	18/09/2020	18/09/2020
					CT314	Electrically powered	E0601	Arrhythmia	A070504	Premature Discharge of Battery	18/09/2020	18/09/2020
									A071102	Failure to Interrogate	18/09/2020	18/09/2020
									A071204	No Pacing	18/09/2020	18/09/2020
							F1901	Additional Surgery	A070504	Premature Discharge of Battery	18/09/2020	18/09/2020
									A071102	Failure to Interrogate	18/09/2020	18/09/2020
									A071204	No Pacing	18/09/2020	18/09/2020
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	18/09/2020	18/09/2020
									A071102	Failure to Interrogate	18/09/2020	18/09/2020
									A071204	No Pacing	18/09/2020	18/09/2020
					CT321	Long-term surgical invasive	E0601	Arrhythmia	A070504	Premature Discharge of Battery	18/09/2020	18/09/2020
									A071102	Failure to Interrogate	18/09/2020	18/09/2020
									A071204	No Pacing	18/09/2020	18/09/2020
							F1901	Additional Surgery	A070504	Premature Discharge of Battery	18/09/2020	18/09/2020
									A071102	Failure to Interrogate	18/09/2020	18/09/2020
									A071204	No Pacing	18/09/2020	18/09/2020
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	18/09/2020	18/09/2020
									A071102	Failure to Interrogate	18/09/2020	18/09/2020
									A071204	No Pacing	18/09/2020	18/09/2020
					CT449	Generators and associated devices	E0601	Arrhythmia	A070504	Premature Discharge of Battery	18/09/2020	18/09/2020
									A071102	Failure to Interrogate	18/09/2020	18/09/2020
									A071204	No Pacing	18/09/2020	18/09/2020
							F1901	Additional Surgery	A070504	Premature Discharge of Battery	18/09/2020	18/09/2020
									A071102	Failure to Interrogate	18/09/2020	18/09/2020
									A071204	No Pacing	18/09/2020	18/09/2020
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	18/09/2020	18/09/2020
									A071102	Failure to Interrogate	18/09/2020	18/09/2020
									A071204	No Pacing	18/09/2020	18/09/2020
					CT981	Single use	E0601	Arrhythmia	A070504	Premature Discharge of Battery	18/09/2020	18/09/2020
									A071102	Failure to Interrogate	18/09/2020	18/09/2020
									A071204	No Pacing	18/09/2020	18/09/2020
							F1901	Additional Surgery	A070504	Premature Discharge of Battery	18/09/2020	18/09/2020
									A071102	Failure to Interrogate	18/09/2020	18/09/2020
									A071204	No Pacing	18/09/2020	18/09/2020
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	18/09/2020	18/09/2020
									A071102	Failure to Interrogate	18/09/2020	18/09/2020
									A071204	No Pacing	18/09/2020	18/09/2020
					CT983	Surgical invasive	E0601	Arrhythmia	A070504	Premature Discharge of Battery	18/09/2020	18/09/2020
									A071102	Failure to Interrogate	18/09/2020	18/09/2020
									A071204	No Pacing	18/09/2020	18/09/2020
							F1901	Additional Surgery	A070504	Premature Discharge of Battery	18/09/2020	18/09/2020
									A071102	Failure to Interrogate	18/09/2020	18/09/2020
									A071204	No Pacing	18/09/2020	18/09/2020
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	18/09/2020	18/09/2020
									A071102	Failure to Interrogate	18/09/2020	18/09/2020
									A071204	No Pacing	18/09/2020	18/09/2020
10059	2020	All other reportable incidents	47265	Dual-chamber implantable pacemaker, rate-responsive	CT141	Electrodes/Leads and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		

					CT145	Cardiovascular devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT233	Surgical	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT2406	Implantable	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT2407	Active implantable	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT2727	Active implant devices/ancillaries	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT2750	Implantable cardiac electrical stimulators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT2985	Active	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT3001	Implantable metal	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT3002	Implantable synthetic polymer	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT3006	Implantable inorganic material	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT314	Electrically powered	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT321	Long-term surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT449	Generators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT981	Single use	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT983	Surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
10060	2019	All other reportable incidents	36241	Pacemaker/defibrillator lead	CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A072201	High Impedance		
10061	2019	All other reportable incidents	47270	Cardiac resynchronization therapy implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
10062	2019	All other reportable incidents	35223	Endocardial/interventricular septal pacing lead	CT1684	Cardiac pacemakers and associated devices	F1905	Device Revision or Replacement	A070101	Failure to Capture	11/10/2019	
10063	2019	All other reportable incidents	35223	Endocardial/interventricular septal pacing lead	CT1684	Cardiac pacemakers and associated devices	F1905	Device Revision or Replacement	A070909	Oversensing	11/10/2019	
10064	2020	All other reportable incidents	37265	Dual-chamber implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	29/09/2020	29/09/2020
							F1901	Additional Surgery	A070504	Premature Discharge of Battery	29/09/2020	29/09/2020
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	29/09/2020	29/09/2020
10065	2020	All other reportable incidents	47265	Dual-chamber implantable pacemaker, rate-responsive	CT141	Electrodes/Leads and associated devices	E0112	Dizziness	A070101	Failure to Capture	16/09/2020	16/09/2020
									A070909	Oversensing	16/09/2020	16/09/2020
									A0712	Pacing Problem	16/09/2020	16/09/2020
							F08	Hospitalization or Prolonged Hospitalization	A070101	Failure to Capture	16/09/2020	16/09/2020
									A070909	Oversensing	16/09/2020	16/09/2020
									A0712	Pacing Problem	16/09/2020	16/09/2020
							F1905	Device Revision or Replacement	A070101	Failure to Capture	16/09/2020	16/09/2020
									A070909	Oversensing	16/09/2020	16/09/2020
									A0712	Pacing Problem	16/09/2020	16/09/2020
					CT145	Cardiovascular devices	E0112	Dizziness	A070101	Failure to Capture	16/09/2020	16/09/2020
									A070909	Oversensing	16/09/2020	16/09/2020
									A0712	Pacing Problem	16/09/2020	16/09/2020
							F08	Hospitalization or Prolonged Hospitalization	A070101	Failure to Capture	16/09/2020	16/09/2020
									A070909	Oversensing	16/09/2020	16/09/2020
									A0712	Pacing Problem	16/09/2020	16/09/2020
							F1905	Device Revision or Replacement	A070101	Failure to Capture	16/09/2020	16/09/2020
									A070909	Oversensing	16/09/2020	16/09/2020
									A0712	Pacing Problem	16/09/2020	16/09/2020
					CT1684	Cardiac pacemakers and associated devices	E0112	Dizziness	A070101	Failure to Capture	16/09/2020	16/09/2020
									A070909	Oversensing	16/09/2020	16/09/2020
									A0712	Pacing Problem	16/09/2020	16/09/2020
							F08	Hospitalization or Prolonged Hospitalization	A070101	Failure to Capture	16/09/2020	16/09/2020
									A070909	Oversensing	16/09/2020	16/09/2020
									A0712	Pacing Problem	16/09/2020	16/09/2020
							F1905	Device Revision or Replacement	A070101	Failure to Capture	16/09/2020	16/09/2020
									A070909	Oversensing	16/09/2020	16/09/2020
									A0712	Pacing Problem	16/09/2020	16/09/2020
					CT1685	Cardiac pacemakers	E0112	Dizziness	A070101	Failure to Capture	16/09/2020	16/09/2020
									A070909	Oversensing	16/09/2020	16/09/2020
									A0712	Pacing Problem	16/09/2020	16/09/2020
							F08	Hospitalization or Prolonged Hospitalization	A070101	Failure to Capture	16/09/2020	16/09/2020
									A070909	Oversensing	16/09/2020	16/09/2020
									A0712	Pacing Problem	16/09/2020	16/09/2020
							F1905	Device Revision or Replacement	A070101	Failure to Capture	16/09/2020	16/09/2020
									A070909	Oversensing	16/09/2020	16/09/2020
									A0712	Pacing Problem	16/09/2020	16/09/2020
					CT233	Surgical	E0112	Dizziness	A070101	Failure to Capture	16/09/2020	16/09/2020
									A070909	Oversensing	16/09/2020	16/09/2020
									A0712	Pacing Problem	16/09/2020	16/09/2020
							F08	Hospitalization or Prolonged Hospitalization	A070101	Failure to Capture	16/09/2020	16/09/2020

								A070909	Oversensing	16/09/2020	16/09/2020
								A0712	Pacing Problem	16/09/2020	16/09/2020
							F1905	Device Revision or Replacement	A070101	Failure to Capture	16/09/2020	16/09/2020
									A070909	Oversensing	16/09/2020	16/09/2020
									A0712	Pacing Problem	16/09/2020	16/09/2020
					CT2406	Implantable	E0112	Dizziness	A070101	Failure to Capture	16/09/2020	16/09/2020
									A070909	Oversensing	16/09/2020	16/09/2020
									A0712	Pacing Problem	16/09/2020	16/09/2020
							F08	Hospitalization or Prolonged Hospitalization	A070101	Failure to Capture	16/09/2020	16/09/2020
									A070909	Oversensing	16/09/2020	16/09/2020
									A0712	Pacing Problem	16/09/2020	16/09/2020
							F1905	Device Revision or Replacement	A070101	Failure to Capture	16/09/2020	16/09/2020
									A070909	Oversensing	16/09/2020	16/09/2020
									A0712	Pacing Problem	16/09/2020	16/09/2020
					CT2407	Active implantable	E0112	Dizziness	A070101	Failure to Capture	16/09/2020	16/09/2020
									A070909	Oversensing	16/09/2020	16/09/2020
									A0712	Pacing Problem	16/09/2020	16/09/2020
							F08	Hospitalization or Prolonged Hospitalization	A070101	Failure to Capture	16/09/2020	16/09/2020
									A070909	Oversensing	16/09/2020	16/09/2020
									A0712	Pacing Problem	16/09/2020	16/09/2020
							F1905	Device Revision or Replacement	A070101	Failure to Capture	16/09/2020	16/09/2020
									A070909	Oversensing	16/09/2020	16/09/2020
									A0712	Pacing Problem	16/09/2020	16/09/2020
					CT257	Implantable cardiac pacemakers	E0112	Dizziness	A070101	Failure to Capture	16/09/2020	16/09/2020
									A070909	Oversensing	16/09/2020	16/09/2020
									A0712	Pacing Problem	16/09/2020	16/09/2020
							F08	Hospitalization or Prolonged Hospitalization	A070101	Failure to Capture	16/09/2020	16/09/2020
									A070909	Oversensing	16/09/2020	16/09/2020
									A0712	Pacing Problem	16/09/2020	16/09/2020
							F1905	Device Revision or Replacement	A070101	Failure to Capture	16/09/2020	16/09/2020
									A070909	Oversensing	16/09/2020	16/09/2020
									A0712	Pacing Problem	16/09/2020	16/09/2020
					CT2727	Active implant devices/ancillaries	E0112	Dizziness	A070101	Failure to Capture	16/09/2020	16/09/2020
									A070909	Oversensing	16/09/2020	16/09/2020
									A0712	Pacing Problem	16/09/2020	16/09/2020
							F08	Hospitalization or Prolonged Hospitalization	A070101	Failure to Capture	16/09/2020	16/09/2020
									A070909	Oversensing	16/09/2020	16/09/2020
									A0712	Pacing Problem	16/09/2020	16/09/2020
							F1905	Device Revision or Replacement	A070101	Failure to Capture	16/09/2020	16/09/2020
									A070909	Oversensing	16/09/2020	16/09/2020
									A0712	Pacing Problem	16/09/2020	16/09/2020
					CT2750	Implantable cardiac electrical stimulators and associated devices	E0112	Dizziness	A070101	Failure to Capture	16/09/2020	16/09/2020
									A070909	Oversensing	16/09/2020	16/09/2020
									A0712	Pacing Problem	16/09/2020	16/09/2020
							F08	Hospitalization or Prolonged Hospitalization	A070101	Failure to Capture	16/09/2020	16/09/2020
									A070909	Oversensing	16/09/2020	16/09/2020
									A0712	Pacing Problem	16/09/2020	16/09/2020
							F1905	Device Revision or Replacement	A070101	Failure to Capture	16/09/2020	16/09/2020
									A070909	Oversensing	16/09/2020	16/09/2020
									A0712	Pacing Problem	16/09/2020	16/09/2020
					CT2866	Implantable cardiac electrical stimulators	E0112	Dizziness	A070101	Failure to Capture	16/09/2020	16/09/2020
									A070909	Oversensing	16/09/2020	16/09/2020
									A0712	Pacing Problem	16/09/2020	16/09/2020
							F08	Hospitalization or Prolonged Hospitalization	A070101	Failure to Capture	16/09/2020	16/09/2020
									A070909	Oversensing	16/09/2020	16/09/2020
									A0712	Pacing Problem	16/09/2020	16/09/2020
							F1905	Device Revision or Replacement	A070101	Failure to Capture	16/09/2020	16/09/2020
									A070909	Oversensing	16/09/2020	16/09/2020
									A0712	Pacing Problem	16/09/2020	16/09/2020
					CT2985	Active	E0112	Dizziness	A070101	Failure to Capture	16/09/2020	16/09/2020
									A070909	Oversensing	16/09/2020	16/09/2020
									A0712	Pacing Problem	16/09/2020	16/09/2020
							F08	Hospitalization or Prolonged Hospitalization	A070101	Failure to Capture	16/09/2020	16/09/2020
									A070909	Oversensing	16/09/2020	16/09/2020
									A0712	Pacing Problem	16/09/2020	16/09/2020
							F1905	Device Revision or Replacement	A070101	Failure to Capture	16/09/2020	16/09/2020
									A070909	Oversensing	16/09/2020	16/09/2020
									A0712	Pacing Problem	16/09/2020	16/09/2020
					CT3001	Implantable metal	E0112	Dizziness	A070101	Failure to Capture	16/09/2020	16/09/2020
									A070909	Oversensing	16/09/2020	16/09/2020
									A0712	Pacing Problem	16/09/2020	16/09/2020
							F08	Hospitalization or Prolonged Hospitalization	A070101	Failure to Capture	16/09/2020	16/09/2020
									A070909	Oversensing	16/09/2020	16/09/2020
									A0712	Pacing Problem	16/09/2020	16/09/2020
							F1905	Device Revision or Replacement	A070101	Failure to Capture	16/09/2020	16/09/2020
									A070909	Oversensing	16/09/2020	16/09/2020
									A0712	Pacing Problem	16/09/2020	16/09/2020
					CT3002	Implantable synthetic polymer	E0112	Dizziness	A070101	Failure to Capture	16/09/2020	16/09/2020
									A070909	Oversensing	16/09/2020	16/09/2020
									A0712	Pacing Problem	16/09/2020	16/09/2020
							F08	Hospitalization or Prolonged Hospitalization	A070101	Failure to Capture	16/09/2020	16/09/2020
									A070909	Oversensing	16/09/2020	16/09/2020
									A0712	Pacing Problem	16/09/2020	16/09/2020
							F1905	Device Revision or Replacement	A070101	Failure to Capture	16/09/2020	16/09/2020
									A070909	Oversensing	16/09/2020	16/09/2020
									A0712	Pacing Problem	16/09/2020	16/09/2020
					CT3006	Implantable inorganic material	E0112	Dizziness	A070101	Failure to Capture	16/09/2020	16/09/2020
									A070909	Oversensing	16/09/2020	16/09/2020

									A0712	Pacing Problem	16/09/2020	16/09/2020
							F08	Hospitalization or Prolonged Hospitalization	A070101	Failure to Capture	16/09/2020	16/09/2020
									A070909	Oversensing	16/09/2020	16/09/2020
									A0712	Pacing Problem	16/09/2020	16/09/2020
							F1905	Device Revision or Replacement	A070101	Failure to Capture	16/09/2020	16/09/2020
									A070909	Oversensing	16/09/2020	16/09/2020
									A0712	Pacing Problem	16/09/2020	16/09/2020
					CT314	Electrically powered	E0112	Dizziness	A070101	Failure to Capture	16/09/2020	16/09/2020
									A070909	Oversensing	16/09/2020	16/09/2020
									A0712	Pacing Problem	16/09/2020	16/09/2020
							F08	Hospitalization or Prolonged Hospitalization	A070101	Failure to Capture	16/09/2020	16/09/2020
									A070909	Oversensing	16/09/2020	16/09/2020
									A0712	Pacing Problem	16/09/2020	16/09/2020
							F1905	Device Revision or Replacement	A070101	Failure to Capture	16/09/2020	16/09/2020
									A070909	Oversensing	16/09/2020	16/09/2020
									A0712	Pacing Problem	16/09/2020	16/09/2020
					CT321	Long-term surgical invasive	E0112	Dizziness	A070101	Failure to Capture	16/09/2020	16/09/2020
									A070909	Oversensing	16/09/2020	16/09/2020
									A0712	Pacing Problem	16/09/2020	16/09/2020
							F08	Hospitalization or Prolonged Hospitalization	A070101	Failure to Capture	16/09/2020	16/09/2020
									A070909	Oversensing	16/09/2020	16/09/2020
									A0712	Pacing Problem	16/09/2020	16/09/2020
							F1905	Device Revision or Replacement	A070101	Failure to Capture	16/09/2020	16/09/2020
									A070909	Oversensing	16/09/2020	16/09/2020
									A0712	Pacing Problem	16/09/2020	16/09/2020
					CT449	Generators and associated devices	E0112	Dizziness	A070101	Failure to Capture	16/09/2020	16/09/2020
									A070909	Oversensing	16/09/2020	16/09/2020
									A0712	Pacing Problem	16/09/2020	16/09/2020
							F08	Hospitalization or Prolonged Hospitalization	A070101	Failure to Capture	16/09/2020	16/09/2020
									A070909	Oversensing	16/09/2020	16/09/2020
									A0712	Pacing Problem	16/09/2020	16/09/2020
							F1905	Device Revision or Replacement	A070101	Failure to Capture	16/09/2020	16/09/2020
									A070909	Oversensing	16/09/2020	16/09/2020
									A0712	Pacing Problem	16/09/2020	16/09/2020
					CT981	Single use	E0112	Dizziness	A070101	Failure to Capture	16/09/2020	16/09/2020
									A070909	Oversensing	16/09/2020	16/09/2020
									A0712	Pacing Problem	16/09/2020	16/09/2020
							F08	Hospitalization or Prolonged Hospitalization	A070101	Failure to Capture	16/09/2020	16/09/2020
									A070909	Oversensing	16/09/2020	16/09/2020
									A0712	Pacing Problem	16/09/2020	16/09/2020
							F1905	Device Revision or Replacement	A070101	Failure to Capture	16/09/2020	16/09/2020
									A070909	Oversensing	16/09/2020	16/09/2020
									A0712	Pacing Problem	16/09/2020	16/09/2020
					CT983	Surgical invasive	E0112	Dizziness	A070101	Failure to Capture	16/09/2020	16/09/2020
									A070909	Oversensing	16/09/2020	16/09/2020
									A0712	Pacing Problem	16/09/2020	16/09/2020
							F08	Hospitalization or Prolonged Hospitalization	A070101	Failure to Capture	16/09/2020	16/09/2020
									A070909	Oversensing	16/09/2020	16/09/2020
									A0712	Pacing Problem	16/09/2020	16/09/2020
							F1905	Device Revision or Replacement	A070101	Failure to Capture	16/09/2020	16/09/2020
									A070909	Oversensing	16/09/2020	16/09/2020
									A0712	Pacing Problem	16/09/2020	16/09/2020
10066	2020	All other reportable incidents	35852	Single-chamber implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A04	Material Integrity Problem		
							F26	No Health Consequences or Impact	A04	Material Integrity Problem		
10067	2020	All other reportable incidents	47265	Dual-chamber implantable pacemaker, rate-responsive	CT141	Electrodes/Leads and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	19/12/2020	19/12/2020
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	19/12/2020	19/12/2020
							F12	Serious Injury/ Illness/ Impairment	A070504	Premature Discharge of Battery	19/12/2020	19/12/2020
							F19	Surgical Intervention	A070504	Premature Discharge of Battery	19/12/2020	19/12/2020
					CT145	Cardiovascular devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	19/12/2020	19/12/2020
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	19/12/2020	19/12/2020
							F12	Serious Injury/ Illness/ Impairment	A070504	Premature Discharge of Battery	19/12/2020	19/12/2020
							F19	Surgical Intervention	A070504	Premature Discharge of Battery	19/12/2020	19/12/2020
					CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	19/12/2020	19/12/2020
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	19/12/2020	19/12/2020
							F12	Serious Injury/ Illness/ Impairment	A070504	Premature Discharge of Battery	19/12/2020	19/12/2020
							F19	Surgical Intervention	A070504	Premature Discharge of Battery	19/12/2020	19/12/2020
					CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	19/12/2020	19/12/2020
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	19/12/2020	19/12/2020
							F12	Serious Injury/ Illness/ Impairment	A070504	Premature Discharge of Battery	19/12/2020	19/12/2020
							F19	Surgical Intervention	A070504	Premature Discharge of Battery	19/12/2020	19/12/2020
					CT233	Surgical	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	19/12/2020	19/12/2020
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	19/12/2020	19/12/2020
							F12	Serious Injury/ Illness/ Impairment	A070504	Premature Discharge of Battery	19/12/2020	19/12/2020
							F19	Surgical Intervention	A070504	Premature Discharge of Battery	19/12/2020	19/12/2020
					CT2406	Implantable	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	19/12/2020	19/12/2020
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	19/12/2020	19/12/2020
							F12	Serious Injury/ Illness/ Impairment	A070504	Premature Discharge of Battery	19/12/2020	19/12/2020
							F19	Surgical Intervention	A070504	Premature Discharge of Battery	19/12/2020	19/12/2020
					CT2407	Active implantable	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	19/12/2020	19/12/2020
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	19/12/2020	19/12/2020
							F12	Serious Injury/ Illness/ Impairment	A070504	Premature Discharge of Battery	19/12/2020	19/12/2020
							F19	Surgical Intervention	A070504	Premature Discharge of Battery	19/12/2020	19/12/2020
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	19/12/2020	19/12/2020
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	19/12/2020	19/12/2020
							F12	Serious Injury/ Illness/ Impairment	A070504	Premature Discharge of Battery	19/12/2020	19/12/2020
							F19	Surgical Intervention	A070504	Premature Discharge of Battery	19/12/2020	19/12/2020
					CT2727	Active implant devices/ancillaries	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	19/12/2020	19/12/2020
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	19/12/2020	19/12/2020

							F12	Serious Injury/ Illness/ Impairment	A070504	Premature Discharge of Battery	19/12/2020	19/12/2020
							F19	Surgical Intervention	A070504	Premature Discharge of Battery	19/12/2020	19/12/2020
					CT2750	Implantable cardiac electrical stimulators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	19/12/2020	19/12/2020
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	19/12/2020	19/12/2020
							F12	Serious Injury/ Illness/ Impairment	A070504	Premature Discharge of Battery	19/12/2020	19/12/2020
							F19	Surgical Intervention	A070504	Premature Discharge of Battery	19/12/2020	19/12/2020
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	19/12/2020	19/12/2020
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	19/12/2020	19/12/2020
							F12	Serious Injury/ Illness/ Impairment	A070504	Premature Discharge of Battery	19/12/2020	19/12/2020
							F19	Surgical Intervention	A070504	Premature Discharge of Battery	19/12/2020	19/12/2020
					CT2985	Active	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	19/12/2020	19/12/2020
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	19/12/2020	19/12/2020
							F12	Serious Injury/ Illness/ Impairment	A070504	Premature Discharge of Battery	19/12/2020	19/12/2020
							F19	Surgical Intervention	A070504	Premature Discharge of Battery	19/12/2020	19/12/2020
					CT3001	Implantable metal	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	19/12/2020	19/12/2020
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	19/12/2020	19/12/2020
							F12	Serious Injury/ Illness/ Impairment	A070504	Premature Discharge of Battery	19/12/2020	19/12/2020
							F19	Surgical Intervention	A070504	Premature Discharge of Battery	19/12/2020	19/12/2020
					CT3002	Implantable synthetic polymer	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	19/12/2020	19/12/2020
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	19/12/2020	19/12/2020
							F12	Serious Injury/ Illness/ Impairment	A070504	Premature Discharge of Battery	19/12/2020	19/12/2020
							F19	Surgical Intervention	A070504	Premature Discharge of Battery	19/12/2020	19/12/2020
					CT3006	Implantable inorganic material	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	19/12/2020	19/12/2020
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	19/12/2020	19/12/2020
							F12	Serious Injury/ Illness/ Impairment	A070504	Premature Discharge of Battery	19/12/2020	19/12/2020
							F19	Surgical Intervention	A070504	Premature Discharge of Battery	19/12/2020	19/12/2020
					CT314	Electrically powered	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	19/12/2020	19/12/2020
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	19/12/2020	19/12/2020
							F12	Serious Injury/ Illness/ Impairment	A070504	Premature Discharge of Battery	19/12/2020	19/12/2020
							F19	Surgical Intervention	A070504	Premature Discharge of Battery	19/12/2020	19/12/2020
					CT321	Long-term surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	19/12/2020	19/12/2020
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	19/12/2020	19/12/2020
							F12	Serious Injury/ Illness/ Impairment	A070504	Premature Discharge of Battery	19/12/2020	19/12/2020
							F19	Surgical Intervention	A070504	Premature Discharge of Battery	19/12/2020	19/12/2020
					CT449	Generators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	19/12/2020	19/12/2020
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	19/12/2020	19/12/2020
							F12	Serious Injury/ Illness/ Impairment	A070504	Premature Discharge of Battery	19/12/2020	19/12/2020
							F19	Surgical Intervention	A070504	Premature Discharge of Battery	19/12/2020	19/12/2020
					CT981	Single use	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	19/12/2020	19/12/2020
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	19/12/2020	19/12/2020
							F12	Serious Injury/ Illness/ Impairment	A070504	Premature Discharge of Battery	19/12/2020	19/12/2020
							F19	Surgical Intervention	A070504	Premature Discharge of Battery	19/12/2020	19/12/2020
					CT983	Surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	19/12/2020	19/12/2020
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	19/12/2020	19/12/2020
							F12	Serious Injury/ Illness/ Impairment	A070504	Premature Discharge of Battery	19/12/2020	19/12/2020
							F19	Surgical Intervention	A070504	Premature Discharge of Battery	19/12/2020	19/12/2020
10068	2019	All other reportable incidents	47270	Cardiac resynchronization therapy implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A070909	Oversensing		
10069	2019	All other reportable incidents	36241	Pacemaker/defibrillator lead	CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070909	Oversensing		
10070	2020	All other reportable incidents	35852	Single-chamber implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	02/10/2020	02/10/2020
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	02/10/2020	02/10/2020
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery	02/10/2020	02/10/2020
10071	2019	All other reportable incidents	36241	Pacemaker/defibrillator lead	CT1684	Cardiac pacemakers and associated devices	F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A150202	Malposition of Device		
							F14	Prolonged Episode of Care	A150202	Malposition of Device		
							F1905	Device Revision or Replacement	A150202	Malposition of Device		
10072	2019	All other reportable incidents	36241	Pacemaker/defibrillator lead	CT1684	Cardiac pacemakers and associated devices	F1905	Device Revision or Replacement	A23	Use of Device Problem		
10073	2019	All other reportable incidents	35852	Single-chamber implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A070909	Oversensing		
10074	2020	All other reportable incidents	47263	Cardiac resynchronization therapy implantable pacemaker	CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070909	Oversensing		
							F26	No Health Consequences or Impact	A070909	Oversensing		
					CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A070909	Oversensing		
							F26	No Health Consequences or Impact	A070909	Oversensing		
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A070909	Oversensing		
							F26	No Health Consequences or Impact	A070909	Oversensing		
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A070909	Oversensing		
							F26	No Health Consequences or Impact	A070909	Oversensing		
10075	2020	All other reportable incidents	36241	Pacemaker/defibrillator lead	CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A051201	Device Dislodged or Dislocated		
									A070101	Failure to Capture		
							F24	Insufficient Health Impact Information	A051201	Device Dislodged or Dislocated		
									A070101	Failure to Capture		
10076	2019	All other reportable incidents	47265	Dual-chamber implantable pacemaker, rate-responsive	CT141	Electrodes/Leads and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	03/10/2019	
					CT145	Cardiovascular devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	03/10/2019	
					CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	03/10/2019	
					CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	03/10/2019	
					CT233	Surgical	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	03/10/2019	
					CT2406	Implantable	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	03/10/2019	
					CT2407	Active implantable	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	03/10/2019	
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	03/10/2019	
					CT2727	Active implant devices/ancillaries	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	03/10/2019	
					CT2750	Implantable cardiac electrical stimulators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	03/10/2019	
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	03/10/2019	
					CT2985	Active	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	03/10/2019	
					CT3001	Implantable metal	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	03/10/2019	
					CT3002	Implantable synthetic polymer	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	03/10/2019	
					CT3006	Implantable inorganic material	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	03/10/2019	
					CT314	Electrically powered	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	03/10/2019	
					CT321	Long-term surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	03/10/2019	
					CT449	Generators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	03/10/2019	

					CT981	Single use	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	03/10/2019	
					CT983	Surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	03/10/2019	
10077	2019	All other reportable incidents	35223	Endocardial/interventricular septal pacing lead	CT1684	Cardiac pacemakers and associated devices	F1905	Device Revision or Replacement	A0709	Device Sensing Problem	25/10/2019	
10078	2019	All other reportable incidents	35852	Single-chamber implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	24/10/2019	
10079	2019	All other reportable incidents	47267	Single-chamber implantable pacemaker, rate-responsive	CT141	Electrodes/Leads and associated devices	F1905	Device Revision or Replacement	A160601	Premature Elective Replacement Indicator		
					CT145	Cardiovascular devices	F1905	Device Revision or Replacement	A160601	Premature Elective Replacement Indicator		
					CT1684	Cardiac pacemakers and associated devices	F1905	Device Revision or Replacement	A160601	Premature Elective Replacement Indicator		
					CT1685	Cardiac pacemakers	F1905	Device Revision or Replacement	A160601	Premature Elective Replacement Indicator		
					CT233	Surgical	F1905	Device Revision or Replacement	A160601	Premature Elective Replacement Indicator		
					CT2406	Implantable	F1905	Device Revision or Replacement	A160601	Premature Elective Replacement Indicator		
					CT2407	Active implantable	F1905	Device Revision or Replacement	A160601	Premature Elective Replacement Indicator		
					CT257	Implantable cardiac pacemakers	F1905	Device Revision or Replacement	A160601	Premature Elective Replacement Indicator		
					CT2727	Active implant devices/ancillaries	F1905	Device Revision or Replacement	A160601	Premature Elective Replacement Indicator		
					CT2750	Implantable cardiac electrical stimulators and associated devices	F1905	Device Revision or Replacement	A160601	Premature Elective Replacement Indicator		
					CT2866	Implantable cardiac electrical stimulators	F1905	Device Revision or Replacement	A160601	Premature Elective Replacement Indicator		
					CT2985	Active	F1905	Device Revision or Replacement	A160601	Premature Elective Replacement Indicator		
					CT3001	Implantable metal	F1905	Device Revision or Replacement	A160601	Premature Elective Replacement Indicator		
					CT3002	Implantable synthetic polymer	F1905	Device Revision or Replacement	A160601	Premature Elective Replacement Indicator		
					CT3006	Implantable inorganic material	F1905	Device Revision or Replacement	A160601	Premature Elective Replacement Indicator		
					CT314	Electrically powered	F1905	Device Revision or Replacement	A160601	Premature Elective Replacement Indicator		
					CT321	Long-term surgical invasive	F1905	Device Revision or Replacement	A160601	Premature Elective Replacement Indicator		
					CT449	Generators and associated devices	F1905	Device Revision or Replacement	A160601	Premature Elective Replacement Indicator		
					CT981	Single use	F1905	Device Revision or Replacement	A160601	Premature Elective Replacement Indicator		
					CT983	Surgical invasive	F1905	Device Revision or Replacement	A160601	Premature Elective Replacement Indicator		
10080	2019	All other reportable incidents	35852	Single-chamber implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A072201	High Impedance		
10081	2019	All other reportable incidents	36241	Pacemaker/defibrillator lead	CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A010402	Migration		
10082	2020	All other reportable incidents	47265	Dual-chamber implantable pacemaker, rate-responsive	CT141	Electrodes/Leads and associated devices	E0112	Dizziness	A071204	No Pacing		
									A17	Compatibility Problem		
							E0127	Numbness	A071204	No Pacing		
									A17	Compatibility Problem		
							F26	No Health Consequences or Impact	A071204	No Pacing		
									A17	Compatibility Problem		
					CT145	Cardiovascular devices	E0112	Dizziness	A071204	No Pacing		
									A17	Compatibility Problem		
							E0127	Numbness	A071204	No Pacing		
									A17	Compatibility Problem		
							F26	No Health Consequences or Impact	A071204	No Pacing		
									A17	Compatibility Problem		
					CT1684	Cardiac pacemakers and associated devices	E0112	Dizziness	A071204	No Pacing		
									A17	Compatibility Problem		
							E0127	Numbness	A071204	No Pacing		
									A17	Compatibility Problem		
							F26	No Health Consequences or Impact	A071204	No Pacing		
									A17	Compatibility Problem		
					CT1685	Cardiac pacemakers	E0112	Dizziness	A071204	No Pacing		
									A17	Compatibility Problem		
							E0127	Numbness	A071204	No Pacing		
									A17	Compatibility Problem		
							F26	No Health Consequences or Impact	A071204	No Pacing		
									A17	Compatibility Problem		
					CT233	Surgical	E0112	Dizziness	A071204	No Pacing		
									A17	Compatibility Problem		
							E0127	Numbness	A071204	No Pacing		
									A17	Compatibility Problem		
							F26	No Health Consequences or Impact	A071204	No Pacing		
									A17	Compatibility Problem		
					CT2406	Implantable	E0112	Dizziness	A071204	No Pacing		
									A17	Compatibility Problem		
							E0127	Numbness	A071204	No Pacing		
									A17	Compatibility Problem		
							F26	No Health Consequences or Impact	A071204	No Pacing		
									A17	Compatibility Problem		
					CT2407	Active implantable	E0112	Dizziness	A071204	No Pacing		
									A17	Compatibility Problem		
							E0127	Numbness	A071204	No Pacing		
									A17	Compatibility Problem		
							F26	No Health Consequences or Impact	A071204	No Pacing		
									A17	Compatibility Problem		
					CT257	Implantable cardiac pacemakers	E0112	Dizziness	A071204	No Pacing		
									A17	Compatibility Problem		
							E0127	Numbness	A071204	No Pacing		
									A17	Compatibility Problem		
							F26	No Health Consequences or Impact	A071204	No Pacing		
									A17	Compatibility Problem		
					CT2727	Active implant devices/ancillaries	E0112	Dizziness	A071204	No Pacing		
									A17	Compatibility Problem		
							E0127	Numbness	A071204	No Pacing		
									A17	Compatibility Problem		
							F26	No Health Consequences or Impact	A071204	No Pacing		
									A17	Compatibility Problem		
					CT2750	Implantable cardiac electrical stimulators and associated devices	E0112	Dizziness	A071204	No Pacing		
									A17	Compatibility Problem		
							E0127	Numbness	A071204	No Pacing		
									A17	Compatibility Problem		
							F26	No Health Consequences or Impact	A071204	No Pacing		
									A17	Compatibility Problem		

					CT2866	Implantable cardiac electrical stimulators	E0112	Dizziness	A071204	No Pacing		
									A17	Compatibility Problem		
							E0127	Numbness	A071204	No Pacing		
									A17	Compatibility Problem		
							F26	No Health Consequences or Impact	A071204	No Pacing		
									A17	Compatibility Problem		
					CT2985	Active	E0112	Dizziness	A071204	No Pacing		
									A17	Compatibility Problem		
							E0127	Numbness	A071204	No Pacing		
									A17	Compatibility Problem		
							F26	No Health Consequences or Impact	A071204	No Pacing		
									A17	Compatibility Problem		
					CT3001	Implantable metal	E0112	Dizziness	A071204	No Pacing		
									A17	Compatibility Problem		
							E0127	Numbness	A071204	No Pacing		
									A17	Compatibility Problem		
							F26	No Health Consequences or Impact	A071204	No Pacing		
									A17	Compatibility Problem		
					CT3002	Implantable synthetic polymer	E0112	Dizziness	A071204	No Pacing		
									A17	Compatibility Problem		
							E0127	Numbness	A071204	No Pacing		
									A17	Compatibility Problem		
							F26	No Health Consequences or Impact	A071204	No Pacing		
									A17	Compatibility Problem		
					CT3006	Implantable inorganic material	E0112	Dizziness	A071204	No Pacing		
									A17	Compatibility Problem		
							E0127	Numbness	A071204	No Pacing		
									A17	Compatibility Problem		
							F26	No Health Consequences or Impact	A071204	No Pacing		
									A17	Compatibility Problem		
					CT314	Electrically powered	E0112	Dizziness	A071204	No Pacing		
									A17	Compatibility Problem		
							E0127	Numbness	A071204	No Pacing		
									A17	Compatibility Problem		
							F26	No Health Consequences or Impact	A071204	No Pacing		
									A17	Compatibility Problem		
					CT321	Long-term surgical invasive	E0112	Dizziness	A071204	No Pacing		
									A17	Compatibility Problem		
							E0127	Numbness	A071204	No Pacing		
									A17	Compatibility Problem		
							F26	No Health Consequences or Impact	A071204	No Pacing		
									A17	Compatibility Problem		
					CT449	Generators and associated devices	E0112	Dizziness	A071204	No Pacing		
									A17	Compatibility Problem		
							E0127	Numbness	A071204	No Pacing		
									A17	Compatibility Problem		
							F26	No Health Consequences or Impact	A071204	No Pacing		
									A17	Compatibility Problem		
					CT981	Single use	E0112	Dizziness	A071204	No Pacing		
									A17	Compatibility Problem		
							E0127	Numbness	A071204	No Pacing		
									A17	Compatibility Problem		
							F26	No Health Consequences or Impact	A071204	No Pacing		
									A17	Compatibility Problem		
					CT983	Surgical invasive	E0112	Dizziness	A071204	No Pacing		
									A17	Compatibility Problem		
							E0127	Numbness	A071204	No Pacing		
									A17	Compatibility Problem		
							F26	No Health Consequences or Impact	A071204	No Pacing		
									A17	Compatibility Problem		
10083	2020	All other reportable incidents	37265	Dual-chamber implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A070909	Oversensing		
							F26	No Health Consequences or Impact	A070909	Oversensing		
10084	2020	All other reportable incidents	47267	Single-chamber implantable pacemaker, rate-responsive	CT141	Electrodes/Leads and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	21/12/2020	21/12/2020
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	21/12/2020	21/12/2020
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	21/12/2020	21/12/2020
					CT145	Cardiovascular devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	21/12/2020	21/12/2020
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	21/12/2020	21/12/2020
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	21/12/2020	21/12/2020
					CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	21/12/2020	21/12/2020
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	21/12/2020	21/12/2020
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	21/12/2020	21/12/2020
					CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	21/12/2020	21/12/2020
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	21/12/2020	21/12/2020
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	21/12/2020	21/12/2020
					CT233	Surgical	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	21/12/2020	21/12/2020
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	21/12/2020	21/12/2020
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	21/12/2020	21/12/2020
					CT2406	Implantable	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	21/12/2020	21/12/2020
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	21/12/2020	21/12/2020
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	21/12/2020	21/12/2020
					CT2407	Active implantable	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	21/12/2020	21/12/2020
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	21/12/2020	21/12/2020
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	21/12/2020	21/12/2020
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	21/12/2020	21/12/2020
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	21/12/2020	21/12/2020
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	21/12/2020	21/12/2020
					CT2727	Active implant devices/ancillaries	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	21/12/2020	21/12/2020
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	21/12/2020	21/12/2020

							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	21/12/2020	21/12/2020
					CT2750	Implantable cardiac electrical stimulators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	21/12/2020	21/12/2020
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	21/12/2020	21/12/2020
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	21/12/2020	21/12/2020
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	21/12/2020	21/12/2020
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	21/12/2020	21/12/2020
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	21/12/2020	21/12/2020
					CT2985	Active	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	21/12/2020	21/12/2020
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	21/12/2020	21/12/2020
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	21/12/2020	21/12/2020
					CT3001	Implantable metal	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	21/12/2020	21/12/2020
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	21/12/2020	21/12/2020
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	21/12/2020	21/12/2020
					CT3002	Implantable synthetic polymer	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	21/12/2020	21/12/2020
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	21/12/2020	21/12/2020
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	21/12/2020	21/12/2020
					CT3006	Implantable inorganic material	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	21/12/2020	21/12/2020
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	21/12/2020	21/12/2020
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	21/12/2020	21/12/2020
					CT314	Electrically powered	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	21/12/2020	21/12/2020
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	21/12/2020	21/12/2020
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	21/12/2020	21/12/2020
					CT321	Long-term surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	21/12/2020	21/12/2020
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	21/12/2020	21/12/2020
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	21/12/2020	21/12/2020
					CT449	Generators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	21/12/2020	21/12/2020
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	21/12/2020	21/12/2020
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	21/12/2020	21/12/2020
					CT981	Single use	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	21/12/2020	21/12/2020
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	21/12/2020	21/12/2020
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	21/12/2020	21/12/2020
					CT983	Surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	21/12/2020	21/12/2020
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	21/12/2020	21/12/2020
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	21/12/2020	21/12/2020
10085	2020	All other reportable incidents	36241	Pacemaker/defibrillator lead	CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A072201	High Impedance		
							F26	No Health Consequences or Impact	A072201	High Impedance		
10086	2020	All other reportable incidents	35852	Single-chamber implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	13/10/2020	13/10/2020
							F19	Surgical Intervention	A070504	Premature Discharge of Battery	13/10/2020	13/10/2020
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	13/10/2020	13/10/2020
10087	2020	All other reportable incidents	35852	Single-chamber implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E2104	Electric Shock	A071302	Inappropriate/Inadequate Shock/Stimulation		
							E2403	No Clinical Signs, Symptoms or Conditions	A071302	Inappropriate/Inadequate Shock/Stimulation		
							F08	Hospitalization or Prolonged Hospitalization	A071302	Inappropriate/Inadequate Shock/Stimulation		
10088	2019	All other reportable incidents	35223	Endocardial/interventricular septal pacing lead	CT1684	Cardiac pacemakers and associated devices	F1905	Device Revision or Replacement	A070909	Oversensing		
10089	2019	All other reportable incidents	37265	Dual-chamber implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	F1905	Device Revision or Replacement	A160601	Premature Elective Replacement Indicator		
10090	2019	All other reportable incidents	47265	Dual-chamber implantable pacemaker, rate-responsive	CT141	Electrodes/Leads and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	31/10/2019	31/10/2019
					CT145	Cardiovascular devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	31/10/2019	31/10/2019
					CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	31/10/2019	31/10/2019
					CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	31/10/2019	31/10/2019
					CT233	Surgical	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	31/10/2019	31/10/2019
					CT2406	Implantable	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	31/10/2019	31/10/2019
					CT2407	Active implantable	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	31/10/2019	31/10/2019
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	31/10/2019	31/10/2019
					CT2727	Active implant devices/ancillaries	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	31/10/2019	31/10/2019
					CT2750	Implantable cardiac electrical stimulators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	31/10/2019	31/10/2019
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	31/10/2019	31/10/2019
					CT2985	Active	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	31/10/2019	31/10/2019
					CT3001	Implantable metal	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	31/10/2019	31/10/2019
					CT3002	Implantable synthetic polymer	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	31/10/2019	31/10/2019
					CT3006	Implantable inorganic material	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	31/10/2019	31/10/2019
					CT314	Electrically powered	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	31/10/2019	31/10/2019
					CT321	Long-term surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	31/10/2019	31/10/2019
					CT449	Generators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	31/10/2019	31/10/2019
					CT981	Single use	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	31/10/2019	31/10/2019
					CT983	Surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	31/10/2019	31/10/2019
10091	2019	All other reportable incidents	36241	Pacemaker/defibrillator lead	CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070909	Oversensing		
10092	2019	All other reportable incidents	37265	Dual-chamber implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A072201	High Impedance		
10093	2019	All other reportable incidents	47265	Dual-chamber implantable pacemaker, rate-responsive	CT141	Electrodes/Leads and associated devices	F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	16/10/2019	
					CT145	Cardiovascular devices	F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	16/10/2019	
					CT1684	Cardiac pacemakers and associated devices	F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	16/10/2019	
					CT1685	Cardiac pacemakers	F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	16/10/2019	
					CT233	Surgical	F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	16/10/2019	
					CT2406	Implantable	F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	16/10/2019	
					CT2407	Active implantable	F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	16/10/2019	
					CT257	Implantable cardiac pacemakers	F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	16/10/2019	
					CT2727	Active implant devices/ancillaries	F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	16/10/2019	
					CT2750	Implantable cardiac electrical stimulators and associated devices	F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	16/10/2019	
					CT2866	Implantable cardiac electrical stimulators	F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	16/10/2019	
					CT2985	Active	F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	16/10/2019	
					CT3001	Implantable metal	F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	16/10/2019	
					CT3002	Implantable synthetic polymer	F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	16/10/2019	
					CT3006	Implantable inorganic material	F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	16/10/2019	
					CT314	Electrically powered	F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	16/10/2019	
					CT321	Long-term surgical invasive	F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	16/10/2019	
					CT449	Generators and associated devices	F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	16/10/2019	
					CT981	Single use	F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	16/10/2019	

					CT983	Surgical invasive	F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	16/10/2019		
10094	2019	All other reportable incidents	36241	Pacemaker/defibrillator lead	CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A072201	High Impedance			
10095	2019	All other reportable incidents	47270	Cardiac resynchronization therapy implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A072201	High Impedance			
10096	2019	All other reportable incidents	47265	Dual-chamber implantable pacemaker, rate-responsive	CT141	Electrodes/Leads and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	11/10/2019		
					CT145	Cardiovascular devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	11/10/2019		
					CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	11/10/2019		
					CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	11/10/2019		
					CT233	Surgical	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	11/10/2019		
					CT2406	Implantable	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	11/10/2019		
					CT2407	Active implantable	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	11/10/2019		
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	11/10/2019		
					CT2727	Active implant devices/ancillaries	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	11/10/2019		
					CT2750	Implantable cardiac electrical stimulators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	11/10/2019		
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	11/10/2019		
					CT2985	Active	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	11/10/2019		
					CT3001	Implantable metal	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	11/10/2019		
					CT3002	Implantable synthetic polymer	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	11/10/2019		
					CT3006	Implantable inorganic material	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	11/10/2019		
					CT314	Electrically powered	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	11/10/2019		
					CT321	Long-term surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	11/10/2019		
					CT449	Generators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	11/10/2019		
					CT981	Single use	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	11/10/2019		
					CT983	Surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	11/10/2019		
10097	2020	Death	47270	Cardiac resynchronization therapy implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E0601	Arrhythmia	A070910	Undersensing	25/09/2020	25/09/2020	
									A071301	Failure to Deliver Shock/Stimulation	25/09/2020	25/09/2020	
							F02	Death	A070910	Undersensing	25/09/2020	25/09/2020	
									A071301	Failure to Deliver Shock/Stimulation	25/09/2020	25/09/2020	
10098	2020	All other reportable incidents	37265	Dual-chamber implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	30/09/2020	30/09/2020	
							F1901	Additional Surgery	A070504	Premature Discharge of Battery	30/09/2020	30/09/2020	
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	30/09/2020	30/09/2020	
10099	2020	All other reportable incidents	35852	Single-chamber implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery			
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery			
10100	2020	All other reportable incidents	37265	Dual-chamber implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	21/09/2020	21/09/2020	
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	21/09/2020	21/09/2020	
							F12	Serious Injury/ Illness/ Impairment	A070504	Premature Discharge of Battery	21/09/2020	21/09/2020	
							F19	Surgical Intervention	A070504	Premature Discharge of Battery	21/09/2020	21/09/2020	
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	21/09/2020	21/09/2020	
10101	2020	All other reportable incidents	17882	Physiologic-monitoring defibrillation system	CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070803	Failure to Power Up			
							F27	Problem Identified During Non-Clinical Procedure	A070803	Failure to Power Up			
10102	2020	Death	47265	Dual-chamber implantable pacemaker, rate-responsive	CT141	Electrodes/Leads and associated devices	E060103	Atrial Flutter	A070504	Premature Discharge of Battery	01/09/2020	01/09/2020	
							E2401	Insufficient Information	A070504	Premature Discharge of Battery	01/09/2020	01/09/2020	
							F02	Death	A070504	Premature Discharge of Battery	01/09/2020	01/09/2020	
					CT145	Cardiovascular devices	E060103	Atrial Flutter	A070504	Premature Discharge of Battery	01/09/2020	01/09/2020	
							E2401	Insufficient Information	A070504	Premature Discharge of Battery	01/09/2020	01/09/2020	
							F02	Death	A070504	Premature Discharge of Battery	01/09/2020	01/09/2020	
					CT1684	Cardiac pacemakers and associated devices	E060103	Atrial Flutter	A070504	Premature Discharge of Battery	01/09/2020	01/09/2020	
							E2401	Insufficient Information	A070504	Premature Discharge of Battery	01/09/2020	01/09/2020	
							F02	Death	A070504	Premature Discharge of Battery	01/09/2020	01/09/2020	
					CT1685	Cardiac pacemakers	E060103	Atrial Flutter	A070504	Premature Discharge of Battery	01/09/2020	01/09/2020	
							E2401	Insufficient Information	A070504	Premature Discharge of Battery	01/09/2020	01/09/2020	
							F02	Death	A070504	Premature Discharge of Battery	01/09/2020	01/09/2020	
					CT233	Surgical	E060103	Atrial Flutter	A070504	Premature Discharge of Battery	01/09/2020	01/09/2020	
							E2401	Insufficient Information	A070504	Premature Discharge of Battery	01/09/2020	01/09/2020	
							F02	Death	A070504	Premature Discharge of Battery	01/09/2020	01/09/2020	
					CT2406	Implantable	E060103	Atrial Flutter	A070504	Premature Discharge of Battery	01/09/2020	01/09/2020	
							E2401	Insufficient Information	A070504	Premature Discharge of Battery	01/09/2020	01/09/2020	
							F02	Death	A070504	Premature Discharge of Battery	01/09/2020	01/09/2020	
					CT2407	Active implantable	E060103	Atrial Flutter	A070504	Premature Discharge of Battery	01/09/2020	01/09/2020	
							E2401	Insufficient Information	A070504	Premature Discharge of Battery	01/09/2020	01/09/2020	
							F02	Death	A070504	Premature Discharge of Battery	01/09/2020	01/09/2020	
					CT257	Implantable cardiac pacemakers	E060103	Atrial Flutter	A070504	Premature Discharge of Battery	01/09/2020	01/09/2020	
							E2401	Insufficient Information	A070504	Premature Discharge of Battery	01/09/2020	01/09/2020	
							F02	Death	A070504	Premature Discharge of Battery	01/09/2020	01/09/2020	
					CT2727	Active implant devices/ancillaries	E060103	Atrial Flutter	A070504	Premature Discharge of Battery	01/09/2020	01/09/2020	
							E2401	Insufficient Information	A070504	Premature Discharge of Battery	01/09/2020	01/09/2020	
							F02	Death	A070504	Premature Discharge of Battery	01/09/2020	01/09/2020	
					CT2750	Implantable cardiac electrical stimulators and associated devices	E060103	Atrial Flutter	A070504	Premature Discharge of Battery	01/09/2020	01/09/2020	
							E2401	Insufficient Information	A070504	Premature Discharge of Battery	01/09/2020	01/09/2020	
							F02	Death	A070504	Premature Discharge of Battery	01/09/2020	01/09/2020	
					CT2866	Implantable cardiac electrical stimulators	E060103	Atrial Flutter	A070504	Premature Discharge of Battery	01/09/2020	01/09/2020	
							E2401	Insufficient Information	A070504	Premature Discharge of Battery	01/09/2020	01/09/2020	
							F02	Death	A070504	Premature Discharge of Battery	01/09/2020	01/09/2020	
					CT2985	Active	E060103	Atrial Flutter	A070504	Premature Discharge of Battery	01/09/2020	01/09/2020	
							E2401	Insufficient Information	A070504	Premature Discharge of Battery	01/09/2020	01/09/2020	
							F02	Death	A070504	Premature Discharge of Battery	01/09/2020	01/09/2020	
					CT3001	Implantable metal	E060103	Atrial Flutter	A070504	Premature Discharge of Battery	01/09/2020	01/09/2020	
							E2401	Insufficient Information	A070504	Premature Discharge of Battery	01/09/2020	01/09/2020	
							F02	Death	A070504	Premature Discharge of Battery	01/09/2020	01/09/2020	
					CT3002	Implantable synthetic polymer	E060103	Atrial Flutter	A070504	Premature Discharge of Battery	01/09/2020	01/09/2020	
							E2401	Insufficient Information	A070504	Premature Discharge of Battery	01/09/2020	01/09/2020	
							F02	Death	A070504	Premature Discharge of Battery	01/09/2020	01/09/2020	
					CT3006	Implantable inorganic material	E060103	Atrial Flutter	A070504	Premature Discharge of Battery	01/09/2020	01/09/2020	
							E2401	Insufficient Information	A070504	Premature Discharge of Battery	01/09/2020	01/09/2020	
							F02	Death	A070504	Premature Discharge of Battery	01/09/2020	01/09/2020	
					CT314	Electrically powered	E060103	Atrial Flutter	A070504	Premature Discharge of Battery	01/09/2020	01/09/2020	
							E2401	Insufficient Information	A070504	Premature Discharge of Battery	01/09/2020	01/09/2020	

							F02	Death	A070504	Premature Discharge of Battery	01/09/2020	01/09/2020
					CT321	Long-term surgical invasive	E060103	Atrial Flutter	A070504	Premature Discharge of Battery	01/09/2020	01/09/2020
							E2401	Insufficient Information	A070504	Premature Discharge of Battery	01/09/2020	01/09/2020
							F02	Death	A070504	Premature Discharge of Battery	01/09/2020	01/09/2020
					CT449	Generators and associated devices	E060103	Atrial Flutter	A070504	Premature Discharge of Battery	01/09/2020	01/09/2020
							E2401	Insufficient Information	A070504	Premature Discharge of Battery	01/09/2020	01/09/2020
							F02	Death	A070504	Premature Discharge of Battery	01/09/2020	01/09/2020
					CT981	Single use	E060103	Atrial Flutter	A070504	Premature Discharge of Battery	01/09/2020	01/09/2020
							E2401	Insufficient Information	A070504	Premature Discharge of Battery	01/09/2020	01/09/2020
							F02	Death	A070504	Premature Discharge of Battery	01/09/2020	01/09/2020
					CT983	Surgical invasive	E060103	Atrial Flutter	A070504	Premature Discharge of Battery	01/09/2020	01/09/2020
							E2401	Insufficient Information	A070504	Premature Discharge of Battery	01/09/2020	01/09/2020
							F02	Death	A070504	Premature Discharge of Battery	01/09/2020	01/09/2020
10103	2020	All other reportable incidents	37265	Dual-chamber implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	01/12/2020	01/12/2020
							F19	Surgical Intervention	A070504	Premature Discharge of Battery	01/12/2020	01/12/2020
10104	2020	All other reportable incidents	47265	Dual-chamber implantable pacemaker, rate-responsive	CT141	Electrodes/Leads and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A071205	Pacemaker Found in Back-Up Mode	28/07/2020	28/07/2020
							F1901	Additional Surgery	A071205	Pacemaker Found in Back-Up Mode	28/07/2020	28/07/2020
							F1905	Device Revision or Replacement	A071205	Pacemaker Found in Back-Up Mode	28/07/2020	28/07/2020
					CT145	Cardiovascular devices	E2403	No Clinical Signs, Symptoms or Conditions	A071205	Pacemaker Found in Back-Up Mode	28/07/2020	28/07/2020
							F1901	Additional Surgery	A071205	Pacemaker Found in Back-Up Mode	28/07/2020	28/07/2020
							F1905	Device Revision or Replacement	A071205	Pacemaker Found in Back-Up Mode	28/07/2020	28/07/2020
					CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A071205	Pacemaker Found in Back-Up Mode	28/07/2020	28/07/2020
							F1901	Additional Surgery	A071205	Pacemaker Found in Back-Up Mode	28/07/2020	28/07/2020
							F1905	Device Revision or Replacement	A071205	Pacemaker Found in Back-Up Mode	28/07/2020	28/07/2020
					CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A071205	Pacemaker Found in Back-Up Mode	28/07/2020	28/07/2020
							F1901	Additional Surgery	A071205	Pacemaker Found in Back-Up Mode	28/07/2020	28/07/2020
							F1905	Device Revision or Replacement	A071205	Pacemaker Found in Back-Up Mode	28/07/2020	28/07/2020
					CT233	Surgical	E2403	No Clinical Signs, Symptoms or Conditions	A071205	Pacemaker Found in Back-Up Mode	28/07/2020	28/07/2020
							F1901	Additional Surgery	A071205	Pacemaker Found in Back-Up Mode	28/07/2020	28/07/2020
							F1905	Device Revision or Replacement	A071205	Pacemaker Found in Back-Up Mode	28/07/2020	28/07/2020
					CT2406	Implantable	E2403	No Clinical Signs, Symptoms or Conditions	A071205	Pacemaker Found in Back-Up Mode	28/07/2020	28/07/2020
							F1901	Additional Surgery	A071205	Pacemaker Found in Back-Up Mode	28/07/2020	28/07/2020
							F1905	Device Revision or Replacement	A071205	Pacemaker Found in Back-Up Mode	28/07/2020	28/07/2020
					CT2407	Active implantable	E2403	No Clinical Signs, Symptoms or Conditions	A071205	Pacemaker Found in Back-Up Mode	28/07/2020	28/07/2020
							F1901	Additional Surgery	A071205	Pacemaker Found in Back-Up Mode	28/07/2020	28/07/2020
							F1905	Device Revision or Replacement	A071205	Pacemaker Found in Back-Up Mode	28/07/2020	28/07/2020
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A071205	Pacemaker Found in Back-Up Mode	28/07/2020	28/07/2020
							F1901	Additional Surgery	A071205	Pacemaker Found in Back-Up Mode	28/07/2020	28/07/2020
							F1905	Device Revision or Replacement	A071205	Pacemaker Found in Back-Up Mode	28/07/2020	28/07/2020
					CT2727	Active implant devices/ancillaries	E2403	No Clinical Signs, Symptoms or Conditions	A071205	Pacemaker Found in Back-Up Mode	28/07/2020	28/07/2020
							F1901	Additional Surgery	A071205	Pacemaker Found in Back-Up Mode	28/07/2020	28/07/2020
							F1905	Device Revision or Replacement	A071205	Pacemaker Found in Back-Up Mode	28/07/2020	28/07/2020
					CT2750	Implantable cardiac electrical stimulators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A071205	Pacemaker Found in Back-Up Mode	28/07/2020	28/07/2020
							F1901	Additional Surgery	A071205	Pacemaker Found in Back-Up Mode	28/07/2020	28/07/2020
							F1905	Device Revision or Replacement	A071205	Pacemaker Found in Back-Up Mode	28/07/2020	28/07/2020
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A071205	Pacemaker Found in Back-Up Mode	28/07/2020	28/07/2020
							F1901	Additional Surgery	A071205	Pacemaker Found in Back-Up Mode	28/07/2020	28/07/2020
							F1905	Device Revision or Replacement	A071205	Pacemaker Found in Back-Up Mode	28/07/2020	28/07/2020
					CT2985	Active	E2403	No Clinical Signs, Symptoms or Conditions	A071205	Pacemaker Found in Back-Up Mode	28/07/2020	28/07/2020
							F1901	Additional Surgery	A071205	Pacemaker Found in Back-Up Mode	28/07/2020	28/07/2020
							F1905	Device Revision or Replacement	A071205	Pacemaker Found in Back-Up Mode	28/07/2020	28/07/2020
					CT3001	Implantable metal	E2403	No Clinical Signs, Symptoms or Conditions	A071205	Pacemaker Found in Back-Up Mode	28/07/2020	28/07/2020
							F1901	Additional Surgery	A071205	Pacemaker Found in Back-Up Mode	28/07/2020	28/07/2020
							F1905	Device Revision or Replacement	A071205	Pacemaker Found in Back-Up Mode	28/07/2020	28/07/2020
					CT3002	Implantable synthetic polymer	E2403	No Clinical Signs, Symptoms or Conditions	A071205	Pacemaker Found in Back-Up Mode	28/07/2020	28/07/2020
							F1901	Additional Surgery	A071205	Pacemaker Found in Back-Up Mode	28/07/2020	28/07/2020
							F1905	Device Revision or Replacement	A071205	Pacemaker Found in Back-Up Mode	28/07/2020	28/07/2020
					CT3006	Implantable inorganic material	E2403	No Clinical Signs, Symptoms or Conditions	A071205	Pacemaker Found in Back-Up Mode	28/07/2020	28/07/2020
							F1901	Additional Surgery	A071205	Pacemaker Found in Back-Up Mode	28/07/2020	28/07/2020
							F1905	Device Revision or Replacement	A071205	Pacemaker Found in Back-Up Mode	28/07/2020	28/07/2020
					CT314	Electrically powered	E2403	No Clinical Signs, Symptoms or Conditions	A071205	Pacemaker Found in Back-Up Mode	28/07/2020	28/07/2020
							F1901	Additional Surgery	A071205	Pacemaker Found in Back-Up Mode	28/07/2020	28/07/2020
							F1905	Device Revision or Replacement	A071205	Pacemaker Found in Back-Up Mode	28/07/2020	28/07/2020
					CT321	Long-term surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A071205	Pacemaker Found in Back-Up Mode	28/07/2020	28/07/2020
							F1901	Additional Surgery	A071205	Pacemaker Found in Back-Up Mode	28/07/2020	28/07/2020
							F1905	Device Revision or Replacement	A071205	Pacemaker Found in Back-Up Mode	28/07/2020	28/07/2020
					CT449	Generators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A071205	Pacemaker Found in Back-Up Mode	28/07/2020	28/07/2020
							F1901	Additional Surgery	A071205	Pacemaker Found in Back-Up Mode	28/07/2020	28/07/2020
							F1905	Device Revision or Replacement	A071205	Pacemaker Found in Back-Up Mode	28/07/2020	28/07/2020
					CT981	Single use	E2403	No Clinical Signs, Symptoms or Conditions	A071205	Pacemaker Found in Back-Up Mode	28/07/2020	28/07/2020
							F1901	Additional Surgery	A071205	Pacemaker Found in Back-Up Mode	28/07/2020	28/07/2020
							F1905	Device Revision or Replacement	A071205	Pacemaker Found in Back-Up Mode	28/07/2020	28/07/2020
					CT983	Surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A071205	Pacemaker Found in Back-Up Mode	28/07/2020	28/07/2020
							F1901	Additional Surgery	A071205	Pacemaker Found in Back-Up Mode	28/07/2020	28/07/2020
							F1905	Device Revision or Replacement	A071205	Pacemaker Found in Back-Up Mode	28/07/2020	28/07/2020
10105	2020	All other reportable incidents	60789	Intracardiac pacemaker	CT130	Catheters and associated devices	E0604	Cardiac Perforation	A24	No Apparent Device Problem		
							E0619	Pericardial Effusion	A24	No Apparent Device Problem		
							E2114	Perforation	A24	No Apparent Device Problem		
							E2321	Low Blood Pressure/ Hypotension	A24	No Apparent Device Problem		
							F1203	Life Threatening Illness or Injury	A24	No Apparent Device Problem		
					CT145	Cardiovascular devices	E0604	Cardiac Perforation	A24	No Apparent Device Problem		
							E0619	Pericardial Effusion	A24	No Apparent Device Problem		
							E2114	Perforation	A24	No Apparent Device Problem		
							E2321	Low Blood Pressure/ Hypotension	A24	No Apparent Device Problem		
							F1203	Life Threatening Illness or Injury	A24	No Apparent Device Problem		
					CT1684	Cardiac pacemakers and associated devices	E0604	Cardiac Perforation	A24	No Apparent Device Problem		
							E0619	Pericardial Effusion	A24	No Apparent Device Problem		
							E2114	Perforation	A24	No Apparent Device Problem		

							E2321	Low Blood Pressure/ Hypotension	A24	No Apparent Device Problem		
							F1203	Life Threatening Illness or Injury	A24	No Apparent Device Problem		
					CT1685	Cardiac pacemakers	E0604	Cardiac Perforation	A24	No Apparent Device Problem		
							E0619	Pericardial Effusion	A24	No Apparent Device Problem		
							E2114	Perforation	A24	No Apparent Device Problem		
							E2321	Low Blood Pressure/ Hypotension	A24	No Apparent Device Problem		
							F1203	Life Threatening Illness or Injury	A24	No Apparent Device Problem		
					CT233	Surgical	E0604	Cardiac Perforation	A24	No Apparent Device Problem		
							E0619	Pericardial Effusion	A24	No Apparent Device Problem		
							E2114	Perforation	A24	No Apparent Device Problem		
							E2321	Low Blood Pressure/ Hypotension	A24	No Apparent Device Problem		
							F1203	Life Threatening Illness or Injury	A24	No Apparent Device Problem		
					CT2406	Implantable	E0604	Cardiac Perforation	A24	No Apparent Device Problem		
							E0619	Pericardial Effusion	A24	No Apparent Device Problem		
							E2114	Perforation	A24	No Apparent Device Problem		
							E2321	Low Blood Pressure/ Hypotension	A24	No Apparent Device Problem		
							F1203	Life Threatening Illness or Injury	A24	No Apparent Device Problem		
					CT2407	Active implantable	E0604	Cardiac Perforation	A24	No Apparent Device Problem		
							E0619	Pericardial Effusion	A24	No Apparent Device Problem		
							E2114	Perforation	A24	No Apparent Device Problem		
							E2321	Low Blood Pressure/ Hypotension	A24	No Apparent Device Problem		
							F1203	Life Threatening Illness or Injury	A24	No Apparent Device Problem		
					CT257	Implantable cardiac pacemakers	E0604	Cardiac Perforation	A24	No Apparent Device Problem		
							E0619	Pericardial Effusion	A24	No Apparent Device Problem		
							E2114	Perforation	A24	No Apparent Device Problem		
							E2321	Low Blood Pressure/ Hypotension	A24	No Apparent Device Problem		
							F1203	Life Threatening Illness or Injury	A24	No Apparent Device Problem		
					CT2727	Active implant devices/ancillaries	E0604	Cardiac Perforation	A24	No Apparent Device Problem		
							E0619	Pericardial Effusion	A24	No Apparent Device Problem		
							E2114	Perforation	A24	No Apparent Device Problem		
							E2321	Low Blood Pressure/ Hypotension	A24	No Apparent Device Problem		
							F1203	Life Threatening Illness or Injury	A24	No Apparent Device Problem		
					CT2750	Implantable cardiac electrical stimulators and associated devices	E0604	Cardiac Perforation	A24	No Apparent Device Problem		
							E0619	Pericardial Effusion	A24	No Apparent Device Problem		
							E2114	Perforation	A24	No Apparent Device Problem		
							E2321	Low Blood Pressure/ Hypotension	A24	No Apparent Device Problem		
							F1203	Life Threatening Illness or Injury	A24	No Apparent Device Problem		
					CT2818	Catheterization/Cannulation/Fluid path devices	E0604	Cardiac Perforation	A24	No Apparent Device Problem		
							E0619	Pericardial Effusion	A24	No Apparent Device Problem		
							E2114	Perforation	A24	No Apparent Device Problem		
							E2321	Low Blood Pressure/ Hypotension	A24	No Apparent Device Problem		
							F1203	Life Threatening Illness or Injury	A24	No Apparent Device Problem		
					CT2866	Implantable cardiac electrical stimulators	E0604	Cardiac Perforation	A24	No Apparent Device Problem		
							E0619	Pericardial Effusion	A24	No Apparent Device Problem		
							E2114	Perforation	A24	No Apparent Device Problem		
							E2321	Low Blood Pressure/ Hypotension	A24	No Apparent Device Problem		
							F1203	Life Threatening Illness or Injury	A24	No Apparent Device Problem		
					CT2985	Active	E0604	Cardiac Perforation	A24	No Apparent Device Problem		
							E0619	Pericardial Effusion	A24	No Apparent Device Problem		
							E2114	Perforation	A24	No Apparent Device Problem		
							E2321	Low Blood Pressure/ Hypotension	A24	No Apparent Device Problem		
							F1203	Life Threatening Illness or Injury	A24	No Apparent Device Problem		
					CT3001	Implantable metal	E0604	Cardiac Perforation	A24	No Apparent Device Problem		
							E0619	Pericardial Effusion	A24	No Apparent Device Problem		
							E2114	Perforation	A24	No Apparent Device Problem		
							E2321	Low Blood Pressure/ Hypotension	A24	No Apparent Device Problem		
							F1203	Life Threatening Illness or Injury	A24	No Apparent Device Problem		
					CT3002	Implantable synthetic polymer	E0604	Cardiac Perforation	A24	No Apparent Device Problem		
							E0619	Pericardial Effusion	A24	No Apparent Device Problem		
							E2114	Perforation	A24	No Apparent Device Problem		
							E2321	Low Blood Pressure/ Hypotension	A24	No Apparent Device Problem		
							F1203	Life Threatening Illness or Injury	A24	No Apparent Device Problem		
					CT3006	Implantable inorganic material	E0604	Cardiac Perforation	A24	No Apparent Device Problem		
							E0619	Pericardial Effusion	A24	No Apparent Device Problem		
							E2114	Perforation	A24	No Apparent Device Problem		
							E2321	Low Blood Pressure/ Hypotension	A24	No Apparent Device Problem		
							F1203	Life Threatening Illness or Injury	A24	No Apparent Device Problem		
					CT314	Electrically powered	E0604	Cardiac Perforation	A24	No Apparent Device Problem		
							E0619	Pericardial Effusion	A24	No Apparent Device Problem		
							E2114	Perforation	A24	No Apparent Device Problem		
							E2321	Low Blood Pressure/ Hypotension	A24	No Apparent Device Problem		
							F1203	Life Threatening Illness or Injury	A24	No Apparent Device Problem		
					CT321	Long-term surgical invasive	E0604	Cardiac Perforation	A24	No Apparent Device Problem		
							E0619	Pericardial Effusion	A24	No Apparent Device Problem		
							E2114	Perforation	A24	No Apparent Device Problem		
							E2321	Low Blood Pressure/ Hypotension	A24	No Apparent Device Problem		
							F1203	Life Threatening Illness or Injury	A24	No Apparent Device Problem		
					CT981	Single use	E0604	Cardiac Perforation	A24	No Apparent Device Problem		
							E0619	Pericardial Effusion	A24	No Apparent Device Problem		
							E2114	Perforation	A24	No Apparent Device Problem		
							E2321	Low Blood Pressure/ Hypotension	A24	No Apparent Device Problem		
							F1203	Life Threatening Illness or Injury	A24	No Apparent Device Problem		
					CT983	Surgical invasive	E0604	Cardiac Perforation	A24	No Apparent Device Problem		
							E0619	Pericardial Effusion	A24	No Apparent Device Problem		
							E2114	Perforation	A24	No Apparent Device Problem		
							E2321	Low Blood Pressure/ Hypotension	A24	No Apparent Device Problem		
							F1203	Life Threatening Illness or Injury	A24	No Apparent Device Problem		
10106	2020	All other reportable incidents	17882	Physiologic-monitoring defibrillation system	CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070908	Failure to Sense		

							F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A070908	Failure to Sense		
							F26	No Health Consequences or Impact	A070908	Failure to Sense		
10107	2020	All other reportable incidents	36241	Pacemaker/defibrillator lead	CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070909	Oversensing		
							F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A070909	Oversensing		
							F26	No Health Consequences or Impact	A070909	Oversensing		
10108	2020	All other reportable incidents	35852	Single-chamber implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	19/12/2020	19/12/2020
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	19/12/2020	19/12/2020
10109	2020	All other reportable incidents	35852	Single-chamber implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E2104	Electric Shock	A070909	Oversensing		
									A071302	Inappropriate/Inadequate Shock/Stimulation		
									A072201	High Impedance		
									A090801	Signal Artifact/Noise		
							F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A070909	Oversensing		
									A071302	Inappropriate/Inadequate Shock/Stimulation		
									A072201	High Impedance		
									A090801	Signal Artifact/Noise		
							F15	Recognized Device or Procedural Complication	A070909	Oversensing		
									A071302	Inappropriate/Inadequate Shock/Stimulation		
									A072201	High Impedance		
									A090801	Signal Artifact/Noise		
10110	2019	All other reportable incidents	37265	Dual-chamber implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	F21	Unexpected Deterioration	A24	No Apparent Device Problem		
10111	2020	All other reportable incidents	47265	Dual-chamber implantable pacemaker, rate-responsive	CT141	Electrodes/Leads and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	08/12/2020	08/12/2020
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	08/12/2020	08/12/2020
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	08/12/2020	08/12/2020
					CT145	Cardiovascular devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	08/12/2020	08/12/2020
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	08/12/2020	08/12/2020
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	08/12/2020	08/12/2020
					CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	08/12/2020	08/12/2020
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	08/12/2020	08/12/2020
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	08/12/2020	08/12/2020
					CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	08/12/2020	08/12/2020
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	08/12/2020	08/12/2020
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	08/12/2020	08/12/2020
					CT233	Surgical	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	08/12/2020	08/12/2020
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	08/12/2020	08/12/2020
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	08/12/2020	08/12/2020
					CT2406	Implantable	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	08/12/2020	08/12/2020
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	08/12/2020	08/12/2020
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	08/12/2020	08/12/2020
					CT2407	Active implantable	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	08/12/2020	08/12/2020
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	08/12/2020	08/12/2020
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	08/12/2020	08/12/2020
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	08/12/2020	08/12/2020
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	08/12/2020	08/12/2020
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	08/12/2020	08/12/2020
					CT2727	Active implant devices/ancillaries	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	08/12/2020	08/12/2020
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	08/12/2020	08/12/2020
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	08/12/2020	08/12/2020
					CT2750	Implantable cardiac electrical stimulators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	08/12/2020	08/12/2020
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	08/12/2020	08/12/2020
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	08/12/2020	08/12/2020
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	08/12/2020	08/12/2020
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	08/12/2020	08/12/2020
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	08/12/2020	08/12/2020
					CT2985	Active	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	08/12/2020	08/12/2020
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	08/12/2020	08/12/2020
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	08/12/2020	08/12/2020
					CT3001	Implantable metal	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	08/12/2020	08/12/2020
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	08/12/2020	08/12/2020
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	08/12/2020	08/12/2020
					CT3002	Implantable synthetic polymer	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	08/12/2020	08/12/2020
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	08/12/2020	08/12/2020
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	08/12/2020	08/12/2020
					CT3006	Implantable inorganic material	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	08/12/2020	08/12/2020
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	08/12/2020	08/12/2020
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	08/12/2020	08/12/2020
					CT314	Electrically powered	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	08/12/2020	08/12/2020
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	08/12/2020	08/12/2020
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	08/12/2020	08/12/2020
					CT321	Long-term surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	08/12/2020	08/12/2020
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	08/12/2020	08/12/2020
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	08/12/2020	08/12/2020
					CT449	Generators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	08/12/2020	08/12/2020
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	08/12/2020	08/12/2020
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	08/12/2020	08/12/2020
					CT981	Single use	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	08/12/2020	08/12/2020
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	08/12/2020	08/12/2020
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	08/12/2020	08/12/2020
					CT983	Surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	08/12/2020	08/12/2020
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	08/12/2020	08/12/2020
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	08/12/2020	08/12/2020
10112	2019	All other reportable incidents	47265	Dual-chamber implantable pacemaker, rate-responsive	CT141	Electrodes/Leads and associated devices	F1905	Device Revision or Replacement	A160601	Premature Elective Replacement Indicator	22/10/2019	
					CT145	Cardiovascular devices	F1905	Device Revision or Replacement	A160601	Premature Elective Replacement Indicator	22/10/2019	
					CT1684	Cardiac pacemakers and associated devices	F1905	Device Revision or Replacement	A160601	Premature Elective Replacement Indicator	22/10/2019	
					CT1685	Cardiac pacemakers	F1905	Device Revision or Replacement	A160601	Premature Elective Replacement Indicator	22/10/2019	
					CT233	Surgical	F1905	Device Revision or Replacement	A160601	Premature Elective Replacement Indicator	22/10/2019	
					CT2406	Implantable	F1905	Device Revision or Replacement	A160601	Premature Elective Replacement Indicator	22/10/2019	
					CT2407	Active implantable	F1905	Device Revision or Replacement	A160601	Premature Elective Replacement Indicator	22/10/2019	
					CT257	Implantable cardiac pacemakers	F1905	Device Revision or Replacement	A160601	Premature Elective Replacement Indicator	22/10/2019	

					CT2727	Active implant devices/ancillaries	F1905	Device Revision or Replacement	A160601	Premature Elective Replacement Indicator	22/10/2019	
					CT2750	Implantable cardiac electrical stimulators and associated devices	F1905	Device Revision or Replacement	A160601	Premature Elective Replacement Indicator	22/10/2019	
					CT2866	Implantable cardiac electrical stimulators	F1905	Device Revision or Replacement	A160601	Premature Elective Replacement Indicator	22/10/2019	
					CT2985	Active	F1905	Device Revision or Replacement	A160601	Premature Elective Replacement Indicator	22/10/2019	
					CT3001	Implantable metal	F1905	Device Revision or Replacement	A160601	Premature Elective Replacement Indicator	22/10/2019	
					CT3002	Implantable synthetic polymer	F1905	Device Revision or Replacement	A160601	Premature Elective Replacement Indicator	22/10/2019	
					CT3006	Implantable inorganic material	F1905	Device Revision or Replacement	A160601	Premature Elective Replacement Indicator	22/10/2019	
					CT314	Electrically powered	F1905	Device Revision or Replacement	A160601	Premature Elective Replacement Indicator	22/10/2019	
					CT321	Long-term surgical invasive	F1905	Device Revision or Replacement	A160601	Premature Elective Replacement Indicator	22/10/2019	
					CT449	Generators and associated devices	F1905	Device Revision or Replacement	A160601	Premature Elective Replacement Indicator	22/10/2019	
					CT981	Single use	F1905	Device Revision or Replacement	A160601	Premature Elective Replacement Indicator	22/10/2019	
					CT983	Surgical invasive	F1905	Device Revision or Replacement	A160601	Premature Elective Replacement Indicator	22/10/2019	
10113	2019	All other reportable incidents	47265	Dual-chamber implantable pacemaker, rate-responsive	CT141	Electrodes/Leads and associated devices	F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A070504	Premature Discharge of Battery		
					CT145	Cardiovascular devices	F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A070504	Premature Discharge of Battery		
					CT1684	Cardiac pacemakers and associated devices	F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A070504	Premature Discharge of Battery		
					CT1685	Cardiac pacemakers	F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A070504	Premature Discharge of Battery		
					CT233	Surgical	F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A070504	Premature Discharge of Battery		
					CT2406	Implantable	F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A070504	Premature Discharge of Battery		
					CT2407	Active implantable	F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A070504	Premature Discharge of Battery		
					CT257	Implantable cardiac pacemakers	F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A070504	Premature Discharge of Battery		
					CT2727	Active implant devices/ancillaries	F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A070504	Premature Discharge of Battery		
					CT2750	Implantable cardiac electrical stimulators and associated devices	F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A070504	Premature Discharge of Battery		
					CT2866	Implantable cardiac electrical stimulators	F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A070504	Premature Discharge of Battery		
					CT2985	Active	F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A070504	Premature Discharge of Battery		
					CT3001	Implantable metal	F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A070504	Premature Discharge of Battery		
					CT3002	Implantable synthetic polymer	F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A070504	Premature Discharge of Battery		
					CT3006	Implantable inorganic material	F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A070504	Premature Discharge of Battery		
					CT314	Electrically powered	F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A070504	Premature Discharge of Battery		
					CT321	Long-term surgical invasive	F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A070504	Premature Discharge of Battery		
					CT449	Generators and associated devices	F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A070504	Premature Discharge of Battery		
					CT981	Single use	F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A070504	Premature Discharge of Battery		
					CT983	Surgical invasive	F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A070504	Premature Discharge of Battery		
10114	2019	All other reportable incidents	36241	Pacemaker/defibrillator lead	CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A010402	Migration		
10115	2019	All other reportable incidents	47265	Dual-chamber implantable pacemaker, rate-responsive	CT141	Electrodes/Leads and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
									A0709	Device Sensing Problem		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
									A0709	Device Sensing Problem		
					CT145	Cardiovascular devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
									A0709	Device Sensing Problem		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
									A0709	Device Sensing Problem		
					CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
									A0709	Device Sensing Problem		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
									A0709	Device Sensing Problem		
					CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
									A0709	Device Sensing Problem		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
									A0709	Device Sensing Problem		
					CT233	Surgical	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
									A0709	Device Sensing Problem		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
									A0709	Device Sensing Problem		
					CT2406	Implantable	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
									A0709	Device Sensing Problem		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
									A0709	Device Sensing Problem		
					CT2407	Active implantable	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
									A0709	Device Sensing Problem		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
									A0709	Device Sensing Problem		
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
									A0709	Device Sensing Problem		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
									A0709	Device Sensing Problem		
					CT2727	Active implant devices/ancillaries	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
									A0709	Device Sensing Problem		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
									A0709	Device Sensing Problem		
					CT2750	Implantable cardiac electrical stimulators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
									A0709	Device Sensing Problem		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
									A0709	Device Sensing Problem		
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
									A0709	Device Sensing Problem		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
									A0709	Device Sensing Problem		
					CT2985	Active	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
									A0709	Device Sensing Problem		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
									A0709	Device Sensing Problem		
					CT3001	Implantable metal	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
									A0709	Device Sensing Problem		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
									A0709	Device Sensing Problem		
					CT3002	Implantable synthetic polymer	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		

									A0709	Device Sensing Problem		
						F26	No Health Consequences or Impact		A070504	Premature Discharge of Battery		
									A0709	Device Sensing Problem		
					CT3006	Implantable inorganic material	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
									A0709	Device Sensing Problem		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
									A0709	Device Sensing Problem		
					CT314	Electrically powered	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
									A0709	Device Sensing Problem		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
									A0709	Device Sensing Problem		
					CT321	Long-term surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
									A0709	Device Sensing Problem		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
									A0709	Device Sensing Problem		
					CT449	Generators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
									A0709	Device Sensing Problem		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
									A0709	Device Sensing Problem		
					CT981	Single use	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
									A0709	Device Sensing Problem		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
									A0709	Device Sensing Problem		
					CT983	Surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
									A0709	Device Sensing Problem		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
									A0709	Device Sensing Problem		
10116	2019	All other reportable incidents	35852	Single-chamber implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A072201	High Impedance		
10117	2019	All other reportable incidents	36241	Pacemaker/defibrillator lead	CT1684	Cardiac pacemakers and associated devices	E2008	Foreign Body In Patient	A0203	Defective Device	29/10/2019	29/10/2019
									A070102	High Capture Threshold	29/10/2019	29/10/2019
							E2403	No Clinical Signs, Symptoms or Conditions	A0203	Defective Device	29/10/2019	29/10/2019
									A070102	High Capture Threshold	29/10/2019	29/10/2019
							F08	Hospitalization or Prolonged Hospitalization	A0203	Defective Device	29/10/2019	29/10/2019
									A070102	High Capture Threshold	29/10/2019	29/10/2019
							F1905	Device Revision or Replacement	A0203	Defective Device	29/10/2019	29/10/2019
									A070102	High Capture Threshold	29/10/2019	29/10/2019
10118	2019	All other reportable incidents	35852	Single-chamber implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A070909	Oversensing		
									A071302	Inappropriate/Inadequate Shock/Stimulation		
10119	2020	All other reportable incidents	47265	Dual-chamber implantable pacemaker, rate-responsive	CT141	Electrodes/Leads and associated devices	E011903	Syncope/Fainting	A070909	Oversensing		
									A0712	Pacing Problem		
							F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A070909	Oversensing		
									A0712	Pacing Problem		
							F12	Serious Injury/ Illness/ Impairment	A070909	Oversensing		
									A0712	Pacing Problem		
					CT145	Cardiovascular devices	E011903	Syncope/Fainting	A070909	Oversensing		
									A0712	Pacing Problem		
							F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A070909	Oversensing		
									A0712	Pacing Problem		
							F12	Serious Injury/ Illness/ Impairment	A070909	Oversensing		
									A0712	Pacing Problem		
					CT1684	Cardiac pacemakers and associated devices	E011903	Syncope/Fainting	A070909	Oversensing		
									A0712	Pacing Problem		
							F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A070909	Oversensing		
									A0712	Pacing Problem		
							F12	Serious Injury/ Illness/ Impairment	A070909	Oversensing		
									A0712	Pacing Problem		
					CT1685	Cardiac pacemakers	E011903	Syncope/Fainting	A070909	Oversensing		
									A0712	Pacing Problem		
							F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A070909	Oversensing		
									A0712	Pacing Problem		
							F12	Serious Injury/ Illness/ Impairment	A070909	Oversensing		
									A0712	Pacing Problem		
					CT233	Surgical	E011903	Syncope/Fainting	A070909	Oversensing		
									A0712	Pacing Problem		
							F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A070909	Oversensing		
									A0712	Pacing Problem		
							F12	Serious Injury/ Illness/ Impairment	A070909	Oversensing		
									A0712	Pacing Problem		
					CT2406	Implantable	E011903	Syncope/Fainting	A070909	Oversensing		
									A0712	Pacing Problem		
							F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A070909	Oversensing		
									A0712	Pacing Problem		
							F12	Serious Injury/ Illness/ Impairment	A070909	Oversensing		
									A0712	Pacing Problem		
					CT2407	Active implantable	E011903	Syncope/Fainting	A070909	Oversensing		
									A0712	Pacing Problem		
							F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A070909	Oversensing		
									A0712	Pacing Problem		
							F12	Serious Injury/ Illness/ Impairment	A070909	Oversensing		
									A0712	Pacing Problem		
					CT257	Implantable cardiac pacemakers	E011903	Syncope/Fainting	A070909	Oversensing		
									A0712	Pacing Problem		
							F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A070909	Oversensing		
									A0712	Pacing Problem		
							F12	Serious Injury/ Illness/ Impairment	A070909	Oversensing		
									A0712	Pacing Problem		
					CT2727	Active implant devices/ancillaries	E011903	Syncope/Fainting	A070909	Oversensing		
									A0712	Pacing Problem		

							F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A070909	Oversensing			
									A0712	Pacing Problem			
							F12	Serious Injury/ Illness/ Impairment	A070909	Oversensing			
									A0712	Pacing Problem			
					CT2750	Implantable cardiac electrical stimulators and associated devices	E011903	Syncope/Fainting	A070909	Oversensing			
									A0712	Pacing Problem			
							F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A070909	Oversensing			
									A0712	Pacing Problem			
							F12	Serious Injury/ Illness/ Impairment	A070909	Oversensing			
									A0712	Pacing Problem			
					CT2866	Implantable cardiac electrical stimulators	E011903	Syncope/Fainting	A070909	Oversensing			
									A0712	Pacing Problem			
							F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A070909	Oversensing			
									A0712	Pacing Problem			
							F12	Serious Injury/ Illness/ Impairment	A070909	Oversensing			
									A0712	Pacing Problem			
					CT2985	Active	E011903	Syncope/Fainting	A070909	Oversensing			
									A0712	Pacing Problem			
							F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A070909	Oversensing			
									A0712	Pacing Problem			
							F12	Serious Injury/ Illness/ Impairment	A070909	Oversensing			
									A0712	Pacing Problem			
					CT3001	Implantable metal	E011903	Syncope/Fainting	A070909	Oversensing			
									A0712	Pacing Problem			
							F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A070909	Oversensing			
									A0712	Pacing Problem			
							F12	Serious Injury/ Illness/ Impairment	A070909	Oversensing			
									A0712	Pacing Problem			
					CT3002	Implantable synthetic polymer	E011903	Syncope/Fainting	A070909	Oversensing			
									A0712	Pacing Problem			
							F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A070909	Oversensing			
									A0712	Pacing Problem			
							F12	Serious Injury/ Illness/ Impairment	A070909	Oversensing			
									A0712	Pacing Problem			
					CT3006	Implantable inorganic material	E011903	Syncope/Fainting	A070909	Oversensing			
									A0712	Pacing Problem			
							F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A070909	Oversensing			
									A0712	Pacing Problem			
							F12	Serious Injury/ Illness/ Impairment	A070909	Oversensing			
									A0712	Pacing Problem			
					CT314	Electrically powered	E011903	Syncope/Fainting	A070909	Oversensing			
									A0712	Pacing Problem			
							F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A070909	Oversensing			
									A0712	Pacing Problem			
							F12	Serious Injury/ Illness/ Impairment	A070909	Oversensing			
									A0712	Pacing Problem			
					CT321	Long-term surgical invasive	E011903	Syncope/Fainting	A070909	Oversensing			
									A0712	Pacing Problem			
							F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A070909	Oversensing			
									A0712	Pacing Problem			
							F12	Serious Injury/ Illness/ Impairment	A070909	Oversensing			
									A0712	Pacing Problem			
					CT449	Generators and associated devices	E011903	Syncope/Fainting	A070909	Oversensing			
									A0712	Pacing Problem			
							F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A070909	Oversensing			
									A0712	Pacing Problem			
							F12	Serious Injury/ Illness/ Impairment	A070909	Oversensing			
									A0712	Pacing Problem			
					CT981	Single use	E011903	Syncope/Fainting	A070909	Oversensing			
									A0712	Pacing Problem			
							F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A070909	Oversensing			
									A0712	Pacing Problem			
							F12	Serious Injury/ Illness/ Impairment	A070909	Oversensing			
									A0712	Pacing Problem			
					CT983	Surgical invasive	E011903	Syncope/Fainting	A070909	Oversensing			
									A0712	Pacing Problem			
							F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A070909	Oversensing			
									A0712	Pacing Problem			
							F12	Serious Injury/ Illness/ Impairment	A070909	Oversensing			
									A0712	Pacing Problem			
10120	2020	All other reportable incidents	47265	Dual-chamber implantable pacemaker, rate-responsive	CT141	Electrodes/Leads and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	27/10/2020	27/10/2020	
							F19	Surgical Intervention	A070504	Premature Discharge of Battery	27/10/2020	27/10/2020	
					CT145	Cardiovascular devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	27/10/2020	27/10/2020	
							F19	Surgical Intervention	A070504	Premature Discharge of Battery	27/10/2020	27/10/2020	
					CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	27/10/2020	27/10/2020	
							F19	Surgical Intervention	A070504	Premature Discharge of Battery	27/10/2020	27/10/2020	
					CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	27/10/2020	27/10/2020	
							F19	Surgical Intervention	A070504	Premature Discharge of Battery	27/10/2020	27/10/2020	
					CT233	Surgical	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	27/10/2020	27/10/2020	
							F19	Surgical Intervention	A070504	Premature Discharge of Battery	27/10/2020	27/10/2020	
					CT2406	Implantable	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	27/10/2020	27/10/2020	
							F19	Surgical Intervention	A070504	Premature Discharge of Battery	27/10/2020	27/10/2020	
					CT2407	Active implantable	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	27/10/2020	27/10/2020	
							F19	Surgical Intervention	A070504	Premature Discharge of Battery	27/10/2020	27/10/2020	
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	27/10/2020	27/10/2020	
							F19	Surgical Intervention	A070504	Premature Discharge of Battery	27/10/2020	27/10/2020	
					CT2727	Active implant devices/ancillaries	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	27/10/2020	27/10/2020	
							F19	Surgical Intervention	A070504	Premature Discharge of Battery	27/10/2020	27/10/2020	

					CT2750	Implantable cardiac electrical stimulators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	27/10/2020	27/10/2020
							F19	Surgical Intervention	A070504	Premature Discharge of Battery	27/10/2020	27/10/2020
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	27/10/2020	27/10/2020
							F19	Surgical Intervention	A070504	Premature Discharge of Battery	27/10/2020	27/10/2020
					CT2985	Active	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	27/10/2020	27/10/2020
							F19	Surgical Intervention	A070504	Premature Discharge of Battery	27/10/2020	27/10/2020
					CT3001	Implantable metal	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	27/10/2020	27/10/2020
							F19	Surgical Intervention	A070504	Premature Discharge of Battery	27/10/2020	27/10/2020
					CT3002	Implantable synthetic polymer	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	27/10/2020	27/10/2020
							F19	Surgical Intervention	A070504	Premature Discharge of Battery	27/10/2020	27/10/2020
					CT3006	Implantable inorganic material	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	27/10/2020	27/10/2020
							F19	Surgical Intervention	A070504	Premature Discharge of Battery	27/10/2020	27/10/2020
					CT314	Electrically powered	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	27/10/2020	27/10/2020
							F19	Surgical Intervention	A070504	Premature Discharge of Battery	27/10/2020	27/10/2020
					CT321	Long-term surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	27/10/2020	27/10/2020
							F19	Surgical Intervention	A070504	Premature Discharge of Battery	27/10/2020	27/10/2020
					CT449	Generators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	27/10/2020	27/10/2020
							F19	Surgical Intervention	A070504	Premature Discharge of Battery	27/10/2020	27/10/2020
					CT981	Single use	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	27/10/2020	27/10/2020
							F19	Surgical Intervention	A070504	Premature Discharge of Battery	27/10/2020	27/10/2020
					CT983	Surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	27/10/2020	27/10/2020
							F19	Surgical Intervention	A070504	Premature Discharge of Battery	27/10/2020	27/10/2020
10121	2020	All other reportable incidents	47265	Dual-chamber implantable pacemaker, rate-responsive	CT141	Electrodes/Leads and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	13/11/2020	13/11/2020
							F19	Surgical Intervention	A070504	Premature Discharge of Battery	13/11/2020	13/11/2020
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	13/11/2020	13/11/2020
					CT145	Cardiovascular devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	13/11/2020	13/11/2020
							F19	Surgical Intervention	A070504	Premature Discharge of Battery	13/11/2020	13/11/2020
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	13/11/2020	13/11/2020
					CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	13/11/2020	13/11/2020
							F19	Surgical Intervention	A070504	Premature Discharge of Battery	13/11/2020	13/11/2020
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	13/11/2020	13/11/2020
					CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	13/11/2020	13/11/2020
							F19	Surgical Intervention	A070504	Premature Discharge of Battery	13/11/2020	13/11/2020
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	13/11/2020	13/11/2020
					CT233	Surgical	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	13/11/2020	13/11/2020
							F19	Surgical Intervention	A070504	Premature Discharge of Battery	13/11/2020	13/11/2020
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	13/11/2020	13/11/2020
					CT2406	Implantable	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	13/11/2020	13/11/2020
							F19	Surgical Intervention	A070504	Premature Discharge of Battery	13/11/2020	13/11/2020
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	13/11/2020	13/11/2020
					CT2407	Active implantable	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	13/11/2020	13/11/2020
							F19	Surgical Intervention	A070504	Premature Discharge of Battery	13/11/2020	13/11/2020
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	13/11/2020	13/11/2020
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	13/11/2020	13/11/2020
							F19	Surgical Intervention	A070504	Premature Discharge of Battery	13/11/2020	13/11/2020
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	13/11/2020	13/11/2020
					CT2727	Active implant devices/ancillaries	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	13/11/2020	13/11/2020
							F19	Surgical Intervention	A070504	Premature Discharge of Battery	13/11/2020	13/11/2020
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	13/11/2020	13/11/2020
					CT2750	Implantable cardiac electrical stimulators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	13/11/2020	13/11/2020
							F19	Surgical Intervention	A070504	Premature Discharge of Battery	13/11/2020	13/11/2020
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	13/11/2020	13/11/2020
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	13/11/2020	13/11/2020
							F19	Surgical Intervention	A070504	Premature Discharge of Battery	13/11/2020	13/11/2020
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	13/11/2020	13/11/2020
					CT2985	Active	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	13/11/2020	13/11/2020
							F19	Surgical Intervention	A070504	Premature Discharge of Battery	13/11/2020	13/11/2020
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	13/11/2020	13/11/2020
					CT3001	Implantable metal	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	13/11/2020	13/11/2020
							F19	Surgical Intervention	A070504	Premature Discharge of Battery	13/11/2020	13/11/2020
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	13/11/2020	13/11/2020
					CT3002	Implantable synthetic polymer	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	13/11/2020	13/11/2020
							F19	Surgical Intervention	A070504	Premature Discharge of Battery	13/11/2020	13/11/2020
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	13/11/2020	13/11/2020
					CT3006	Implantable inorganic material	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	13/11/2020	13/11/2020
							F19	Surgical Intervention	A070504	Premature Discharge of Battery	13/11/2020	13/11/2020
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	13/11/2020	13/11/2020
					CT314	Electrically powered	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	13/11/2020	13/11/2020
							F19	Surgical Intervention	A070504	Premature Discharge of Battery	13/11/2020	13/11/2020
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	13/11/2020	13/11/2020
					CT321	Long-term surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	13/11/2020	13/11/2020
							F19	Surgical Intervention	A070504	Premature Discharge of Battery	13/11/2020	13/11/2020
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	13/11/2020	13/11/2020
					CT449	Generators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	13/11/2020	13/11/2020
							F19	Surgical Intervention	A070504	Premature Discharge of Battery	13/11/2020	13/11/2020
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	13/11/2020	13/11/2020
					CT981	Single use	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	13/11/2020	13/11/2020
							F19	Surgical Intervention	A070504	Premature Discharge of Battery	13/11/2020	13/11/2020
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	13/11/2020	13/11/2020
					CT983	Surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	13/11/2020	13/11/2020
							F19	Surgical Intervention	A070504	Premature Discharge of Battery	13/11/2020	13/11/2020
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	13/11/2020	13/11/2020
10122	2020	All other reportable incidents	47265	Dual-chamber implantable pacemaker, rate-responsive	CT141	Electrodes/Leads and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	03/11/2020	03/11/2020
							F19	Surgical Intervention	A070504	Premature Discharge of Battery	03/11/2020	03/11/2020
					CT145	Cardiovascular devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	03/11/2020	03/11/2020
							F19	Surgical Intervention	A070504	Premature Discharge of Battery	03/11/2020	03/11/2020
					CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	03/11/2020	03/11/2020
							F19	Surgical Intervention	A070504	Premature Discharge of Battery	03/11/2020	03/11/2020

					CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	03/11/2020	03/11/2020
							F19	Surgical Intervention	A070504	Premature Discharge of Battery	03/11/2020	03/11/2020
					CT233	Surgical	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	03/11/2020	03/11/2020
							F19	Surgical Intervention	A070504	Premature Discharge of Battery	03/11/2020	03/11/2020
					CT2406	Implantable	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	03/11/2020	03/11/2020
							F19	Surgical Intervention	A070504	Premature Discharge of Battery	03/11/2020	03/11/2020
					CT2407	Active implantable	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	03/11/2020	03/11/2020
							F19	Surgical Intervention	A070504	Premature Discharge of Battery	03/11/2020	03/11/2020
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	03/11/2020	03/11/2020
							F19	Surgical Intervention	A070504	Premature Discharge of Battery	03/11/2020	03/11/2020
					CT2727	Active implant devices/ancillaries	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	03/11/2020	03/11/2020
							F19	Surgical Intervention	A070504	Premature Discharge of Battery	03/11/2020	03/11/2020
					CT2750	Implantable cardiac electrical stimulators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	03/11/2020	03/11/2020
							F19	Surgical Intervention	A070504	Premature Discharge of Battery	03/11/2020	03/11/2020
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	03/11/2020	03/11/2020
							F19	Surgical Intervention	A070504	Premature Discharge of Battery	03/11/2020	03/11/2020
					CT2985	Active	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	03/11/2020	03/11/2020
							F19	Surgical Intervention	A070504	Premature Discharge of Battery	03/11/2020	03/11/2020
					CT3001	Implantable metal	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	03/11/2020	03/11/2020
							F19	Surgical Intervention	A070504	Premature Discharge of Battery	03/11/2020	03/11/2020
					CT3002	Implantable synthetic polymer	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	03/11/2020	03/11/2020
							F19	Surgical Intervention	A070504	Premature Discharge of Battery	03/11/2020	03/11/2020
					CT3006	Implantable inorganic material	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	03/11/2020	03/11/2020
							F19	Surgical Intervention	A070504	Premature Discharge of Battery	03/11/2020	03/11/2020
					CT314	Electrically powered	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	03/11/2020	03/11/2020
							F19	Surgical Intervention	A070504	Premature Discharge of Battery	03/11/2020	03/11/2020
					CT321	Long-term surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	03/11/2020	03/11/2020
							F19	Surgical Intervention	A070504	Premature Discharge of Battery	03/11/2020	03/11/2020
					CT449	Generators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	03/11/2020	03/11/2020
							F19	Surgical Intervention	A070504	Premature Discharge of Battery	03/11/2020	03/11/2020
					CT981	Single use	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	03/11/2020	03/11/2020
							F19	Surgical Intervention	A070504	Premature Discharge of Battery	03/11/2020	03/11/2020
					CT983	Surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	03/11/2020	03/11/2020
							F19	Surgical Intervention	A070504	Premature Discharge of Battery	03/11/2020	03/11/2020
10123	2019	All other reportable incidents	17882	Physiologic-monitoring defibrillation system	CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A0712	Pacing Problem		
10124	2019	All other reportable incidents	47265	Dual-chamber implantable pacemaker, rate-responsive	CT141	Electrodes/Leads and associated devices	F1905	Device Revision or Replacement	A070909	Oversensing	31/10/2019	
					CT145	Cardiovascular devices	F1905	Device Revision or Replacement	A070909	Oversensing	31/10/2019	
					CT1684	Cardiac pacemakers and associated devices	F1905	Device Revision or Replacement	A070909	Oversensing	31/10/2019	
					CT1685	Cardiac pacemakers	F1905	Device Revision or Replacement	A070909	Oversensing	31/10/2019	
					CT233	Surgical	F1905	Device Revision or Replacement	A070909	Oversensing	31/10/2019	
					CT2406	Implantable	F1905	Device Revision or Replacement	A070909	Oversensing	31/10/2019	
					CT2407	Active implantable	F1905	Device Revision or Replacement	A070909	Oversensing	31/10/2019	
					CT257	Implantable cardiac pacemakers	F1905	Device Revision or Replacement	A070909	Oversensing	31/10/2019	
					CT2727	Active implant devices/ancillaries	F1905	Device Revision or Replacement	A070909	Oversensing	31/10/2019	
					CT2750	Implantable cardiac electrical stimulators and associated devices	F1905	Device Revision or Replacement	A070909	Oversensing	31/10/2019	
					CT2866	Implantable cardiac electrical stimulators	F1905	Device Revision or Replacement	A070909	Oversensing	31/10/2019	
					CT2985	Active	F1905	Device Revision or Replacement	A070909	Oversensing	31/10/2019	
					CT3001	Implantable metal	F1905	Device Revision or Replacement	A070909	Oversensing	31/10/2019	
					CT3002	Implantable synthetic polymer	F1905	Device Revision or Replacement	A070909	Oversensing	31/10/2019	
					CT3006	Implantable inorganic material	F1905	Device Revision or Replacement	A070909	Oversensing	31/10/2019	
					CT314	Electrically powered	F1905	Device Revision or Replacement	A070909	Oversensing	31/10/2019	
					CT321	Long-term surgical invasive	F1905	Device Revision or Replacement	A070909	Oversensing	31/10/2019	
					CT449	Generators and associated devices	F1905	Device Revision or Replacement	A070909	Oversensing	31/10/2019	
					CT981	Single use	F1905	Device Revision or Replacement	A070909	Oversensing	31/10/2019	
					CT983	Surgical invasive	F1905	Device Revision or Replacement	A070909	Oversensing	31/10/2019	
10125	2019	All other reportable incidents	47270	Cardiac resynchronization therapy implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	F1905	Device Revision or Replacement	A0705	Battery Problem	07/11/2019	
10126	2019	All other reportable incidents	47265	Dual-chamber implantable pacemaker, rate-responsive	CT141	Electrodes/Leads and associated devices	F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	31/10/2019	
					CT145	Cardiovascular devices	F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	31/10/2019	
					CT1684	Cardiac pacemakers and associated devices	F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	31/10/2019	
					CT1685	Cardiac pacemakers	F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	31/10/2019	
					CT233	Surgical	F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	31/10/2019	
					CT2406	Implantable	F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	31/10/2019	
					CT2407	Active implantable	F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	31/10/2019	
					CT257	Implantable cardiac pacemakers	F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	31/10/2019	
					CT2727	Active implant devices/ancillaries	F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	31/10/2019	
					CT2750	Implantable cardiac electrical stimulators and associated devices	F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	31/10/2019	
					CT2866	Implantable cardiac electrical stimulators	F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	31/10/2019	
					CT2985	Active	F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	31/10/2019	
					CT3001	Implantable metal	F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	31/10/2019	
					CT3002	Implantable synthetic polymer	F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	31/10/2019	
					CT3006	Implantable inorganic material	F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	31/10/2019	
					CT314	Electrically powered	F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	31/10/2019	
					CT321	Long-term surgical invasive	F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	31/10/2019	
					CT449	Generators and associated devices	F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	31/10/2019	
					CT981	Single use	F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	31/10/2019	
					CT983	Surgical invasive	F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	31/10/2019	
10127	2019	All other reportable incidents	47265	Dual-chamber implantable pacemaker, rate-responsive	CT141	Electrodes/Leads and associated devices	F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	11/12/2019	11/12/2019
					CT145	Cardiovascular devices	F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	11/12/2019	11/12/2019
					CT1684	Cardiac pacemakers and associated devices	F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	11/12/2019	11/12/2019
					CT1685	Cardiac pacemakers	F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	11/12/2019	11/12/2019
					CT233	Surgical	F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	11/12/2019	11/12/2019
					CT2406	Implantable	F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	11/12/2019	11/12/2019
					CT2407	Active implantable	F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	11/12/2019	11/12/2019
					CT257	Implantable cardiac pacemakers	F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	11/12/2019	11/12/2019
					CT2727	Active implant devices/ancillaries	F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	11/12/2019	11/12/2019

					CT2750	Implantable cardiac electrical stimulators and associated devices	F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	11/12/2019	11/12/2019
					CT2866	Implantable cardiac electrical stimulators	F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	11/12/2019	11/12/2019
					CT2985	Active	F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	11/12/2019	11/12/2019
					CT3001	Implantable metal	F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	11/12/2019	11/12/2019
					CT3002	Implantable synthetic polymer	F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	11/12/2019	11/12/2019
					CT3006	Implantable inorganic material	F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	11/12/2019	11/12/2019
					CT314	Electrically powered	F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	11/12/2019	11/12/2019
					CT321	Long-term surgical invasive	F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	11/12/2019	11/12/2019
					CT449	Generators and associated devices	F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	11/12/2019	11/12/2019
					CT981	Single use	F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	11/12/2019	11/12/2019
					CT983	Surgical invasive	F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	11/12/2019	11/12/2019
10128	2019	All other reportable incidents	36241	Pacemaker/defibrillator lead	CT1684	Cardiac pacemakers and associated devices	E2114	Perforation	A0104	Migration or Expulsion of Device		
10129	2019	All other reportable incidents	47266	Single-chamber implantable pacemaker, fixed-rate/demand	CT141	Electrodes/Leads and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A0711	Interrogation Problem		
					CT145	Cardiovascular devices	E2403	No Clinical Signs, Symptoms or Conditions	A0711	Interrogation Problem		
					CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A0711	Interrogation Problem		
					CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A0711	Interrogation Problem		
					CT233	Surgical	E2403	No Clinical Signs, Symptoms or Conditions	A0711	Interrogation Problem		
					CT2406	Implantable	E2403	No Clinical Signs, Symptoms or Conditions	A0711	Interrogation Problem		
					CT2407	Active implantable	E2403	No Clinical Signs, Symptoms or Conditions	A0711	Interrogation Problem		
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A0711	Interrogation Problem		
					CT2727	Active implant devices/ancillaries	E2403	No Clinical Signs, Symptoms or Conditions	A0711	Interrogation Problem		
					CT2750	Implantable cardiac electrical stimulators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A0711	Interrogation Problem		
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A0711	Interrogation Problem		
					CT2985	Active	E2403	No Clinical Signs, Symptoms or Conditions	A0711	Interrogation Problem		
					CT3001	Implantable metal	E2403	No Clinical Signs, Symptoms or Conditions	A0711	Interrogation Problem		
					CT3002	Implantable synthetic polymer	E2403	No Clinical Signs, Symptoms or Conditions	A0711	Interrogation Problem		
					CT3006	Implantable inorganic material	E2403	No Clinical Signs, Symptoms or Conditions	A0711	Interrogation Problem		
					CT314	Electrically powered	E2403	No Clinical Signs, Symptoms or Conditions	A0711	Interrogation Problem		
					CT321	Long-term surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A0711	Interrogation Problem		
					CT449	Generators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A0711	Interrogation Problem		
					CT981	Single use	E2403	No Clinical Signs, Symptoms or Conditions	A0711	Interrogation Problem		
					CT983	Surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A0711	Interrogation Problem		
10130	2019	All other reportable incidents	35852	Single-chamber implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A070909	Oversensing		
10131	2019	All other reportable incidents	47265	Dual-chamber implantable pacemaker, rate-responsive	CT141	Electrodes/Leads and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A071102	Failure to Interrogate		
					CT145	Cardiovascular devices	E2403	No Clinical Signs, Symptoms or Conditions	A071102	Failure to Interrogate		
					CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A071102	Failure to Interrogate		
					CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A071102	Failure to Interrogate		
					CT233	Surgical	E2403	No Clinical Signs, Symptoms or Conditions	A071102	Failure to Interrogate		
					CT2406	Implantable	E2403	No Clinical Signs, Symptoms or Conditions	A071102	Failure to Interrogate		
					CT2407	Active implantable	E2403	No Clinical Signs, Symptoms or Conditions	A071102	Failure to Interrogate		
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A071102	Failure to Interrogate		
					CT2727	Active implant devices/ancillaries	E2403	No Clinical Signs, Symptoms or Conditions	A071102	Failure to Interrogate		
					CT2750	Implantable cardiac electrical stimulators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A071102	Failure to Interrogate		
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A071102	Failure to Interrogate		
					CT2985	Active	E2403	No Clinical Signs, Symptoms or Conditions	A071102	Failure to Interrogate		
					CT3001	Implantable metal	E2403	No Clinical Signs, Symptoms or Conditions	A071102	Failure to Interrogate		
					CT3002	Implantable synthetic polymer	E2403	No Clinical Signs, Symptoms or Conditions	A071102	Failure to Interrogate		
					CT3006	Implantable inorganic material	E2403	No Clinical Signs, Symptoms or Conditions	A071102	Failure to Interrogate		
					CT314	Electrically powered	E2403	No Clinical Signs, Symptoms or Conditions	A071102	Failure to Interrogate		
					CT321	Long-term surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A071102	Failure to Interrogate		
					CT449	Generators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A071102	Failure to Interrogate		
					CT981	Single use	E2403	No Clinical Signs, Symptoms or Conditions	A071102	Failure to Interrogate		
					CT983	Surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A071102	Failure to Interrogate		
10132	2020	All other reportable incidents	47265	Dual-chamber implantable pacemaker, rate-responsive	CT141	Electrodes/Leads and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070101	Failure to Capture	17/10/2020	17/10/2020
									A070504	Premature Discharge of Battery	17/10/2020	17/10/2020
									A072201	High Impedance	17/10/2020	17/10/2020
							F19	Surgical Intervention	A070101	Failure to Capture	17/10/2020	17/10/2020
									A070504	Premature Discharge of Battery	17/10/2020	17/10/2020
									A072201	High Impedance	17/10/2020	17/10/2020
					CT145	Cardiovascular devices	E2403	No Clinical Signs, Symptoms or Conditions	A070101	Failure to Capture	17/10/2020	17/10/2020
									A070504	Premature Discharge of Battery	17/10/2020	17/10/2020
									A072201	High Impedance	17/10/2020	17/10/2020
							F19	Surgical Intervention	A070101	Failure to Capture	17/10/2020	17/10/2020
									A070504	Premature Discharge of Battery	17/10/2020	17/10/2020
									A072201	High Impedance	17/10/2020	17/10/2020
					CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070101	Failure to Capture	17/10/2020	17/10/2020
									A070504	Premature Discharge of Battery	17/10/2020	17/10/2020
									A072201	High Impedance	17/10/2020	17/10/2020
							F19	Surgical Intervention	A070101	Failure to Capture	17/10/2020	17/10/2020
									A070504	Premature Discharge of Battery	17/10/2020	17/10/2020
									A072201	High Impedance	17/10/2020	17/10/2020
					CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A070101	Failure to Capture	17/10/2020	17/10/2020
									A070504	Premature Discharge of Battery	17/10/2020	17/10/2020
									A072201	High Impedance	17/10/2020	17/10/2020
							F19	Surgical Intervention	A070101	Failure to Capture	17/10/2020	17/10/2020
									A070504	Premature Discharge of Battery	17/10/2020	17/10/2020
									A072201	High Impedance	17/10/2020	17/10/2020
					CT233	Surgical	E2403	No Clinical Signs, Symptoms or Conditions	A070101	Failure to Capture	17/10/2020	17/10/2020
									A070504	Premature Discharge of Battery	17/10/2020	17/10/2020
									A072201	High Impedance	17/10/2020	17/10/2020
							F19	Surgical Intervention	A070101	Failure to Capture	17/10/2020	17/10/2020
									A070504	Premature Discharge of Battery	17/10/2020	17/10/2020
									A072201	High Impedance	17/10/2020	17/10/2020
					CT2406	Implantable	E2403	No Clinical Signs, Symptoms or Conditions	A070101	Failure to Capture	17/10/2020	17/10/2020
									A070504	Premature Discharge of Battery	17/10/2020	17/10/2020

									A072201	High Impedance	17/10/2020	17/10/2020
							F19	Surgical Intervention	A070101	Failure to Capture	17/10/2020	17/10/2020
									A070504	Premature Discharge of Battery	17/10/2020	17/10/2020
									A072201	High Impedance	17/10/2020	17/10/2020
					CT2407	Active implantable	E2403	No Clinical Signs, Symptoms or Conditions	A070101	Failure to Capture	17/10/2020	17/10/2020
									A070504	Premature Discharge of Battery	17/10/2020	17/10/2020
									A072201	High Impedance	17/10/2020	17/10/2020
							F19	Surgical Intervention	A070101	Failure to Capture	17/10/2020	17/10/2020
									A070504	Premature Discharge of Battery	17/10/2020	17/10/2020
									A072201	High Impedance	17/10/2020	17/10/2020
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A070101	Failure to Capture	17/10/2020	17/10/2020
									A070504	Premature Discharge of Battery	17/10/2020	17/10/2020
									A072201	High Impedance	17/10/2020	17/10/2020
							F19	Surgical Intervention	A070101	Failure to Capture	17/10/2020	17/10/2020
									A070504	Premature Discharge of Battery	17/10/2020	17/10/2020
									A072201	High Impedance	17/10/2020	17/10/2020
					CT2727	Active implant devices/ancillaries	E2403	No Clinical Signs, Symptoms or Conditions	A070101	Failure to Capture	17/10/2020	17/10/2020
									A070504	Premature Discharge of Battery	17/10/2020	17/10/2020
									A072201	High Impedance	17/10/2020	17/10/2020
							F19	Surgical Intervention	A070101	Failure to Capture	17/10/2020	17/10/2020
									A070504	Premature Discharge of Battery	17/10/2020	17/10/2020
									A072201	High Impedance	17/10/2020	17/10/2020
					CT2750	Implantable cardiac electrical stimulators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070101	Failure to Capture	17/10/2020	17/10/2020
									A070504	Premature Discharge of Battery	17/10/2020	17/10/2020
									A072201	High Impedance	17/10/2020	17/10/2020
							F19	Surgical Intervention	A070101	Failure to Capture	17/10/2020	17/10/2020
									A070504	Premature Discharge of Battery	17/10/2020	17/10/2020
									A072201	High Impedance	17/10/2020	17/10/2020
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A070101	Failure to Capture	17/10/2020	17/10/2020
									A070504	Premature Discharge of Battery	17/10/2020	17/10/2020
									A072201	High Impedance	17/10/2020	17/10/2020
							F19	Surgical Intervention	A070101	Failure to Capture	17/10/2020	17/10/2020
									A070504	Premature Discharge of Battery	17/10/2020	17/10/2020
									A072201	High Impedance	17/10/2020	17/10/2020
					CT2985	Active	E2403	No Clinical Signs, Symptoms or Conditions	A070101	Failure to Capture	17/10/2020	17/10/2020
									A070504	Premature Discharge of Battery	17/10/2020	17/10/2020
									A072201	High Impedance	17/10/2020	17/10/2020
							F19	Surgical Intervention	A070101	Failure to Capture	17/10/2020	17/10/2020
									A070504	Premature Discharge of Battery	17/10/2020	17/10/2020
									A072201	High Impedance	17/10/2020	17/10/2020
					CT3001	Implantable metal	E2403	No Clinical Signs, Symptoms or Conditions	A070101	Failure to Capture	17/10/2020	17/10/2020
									A070504	Premature Discharge of Battery	17/10/2020	17/10/2020
									A072201	High Impedance	17/10/2020	17/10/2020
							F19	Surgical Intervention	A070101	Failure to Capture	17/10/2020	17/10/2020
									A070504	Premature Discharge of Battery	17/10/2020	17/10/2020
									A072201	High Impedance	17/10/2020	17/10/2020
					CT3002	Implantable synthetic polymer	E2403	No Clinical Signs, Symptoms or Conditions	A070101	Failure to Capture	17/10/2020	17/10/2020
									A070504	Premature Discharge of Battery	17/10/2020	17/10/2020
									A072201	High Impedance	17/10/2020	17/10/2020
							F19	Surgical Intervention	A070101	Failure to Capture	17/10/2020	17/10/2020
									A070504	Premature Discharge of Battery	17/10/2020	17/10/2020
									A072201	High Impedance	17/10/2020	17/10/2020
					CT3006	Implantable inorganic material	E2403	No Clinical Signs, Symptoms or Conditions	A070101	Failure to Capture	17/10/2020	17/10/2020
									A070504	Premature Discharge of Battery	17/10/2020	17/10/2020
									A072201	High Impedance	17/10/2020	17/10/2020
							F19	Surgical Intervention	A070101	Failure to Capture	17/10/2020	17/10/2020
									A070504	Premature Discharge of Battery	17/10/2020	17/10/2020
									A072201	High Impedance	17/10/2020	17/10/2020
					CT314	Electrically powered	E2403	No Clinical Signs, Symptoms or Conditions	A070101	Failure to Capture	17/10/2020	17/10/2020
									A070504	Premature Discharge of Battery	17/10/2020	17/10/2020
									A072201	High Impedance	17/10/2020	17/10/2020
							F19	Surgical Intervention	A070101	Failure to Capture	17/10/2020	17/10/2020
									A070504	Premature Discharge of Battery	17/10/2020	17/10/2020
									A072201	High Impedance	17/10/2020	17/10/2020
					CT321	Long-term surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A070101	Failure to Capture	17/10/2020	17/10/2020
									A070504	Premature Discharge of Battery	17/10/2020	17/10/2020
									A072201	High Impedance	17/10/2020	17/10/2020
							F19	Surgical Intervention	A070101	Failure to Capture	17/10/2020	17/10/2020
									A070504	Premature Discharge of Battery	17/10/2020	17/10/2020
									A072201	High Impedance	17/10/2020	17/10/2020
					CT449	Generators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070101	Failure to Capture	17/10/2020	17/10/2020
									A070504	Premature Discharge of Battery	17/10/2020	17/10/2020
									A072201	High Impedance	17/10/2020	17/10/2020
							F19	Surgical Intervention	A070101	Failure to Capture	17/10/2020	17/10/2020
									A070504	Premature Discharge of Battery	17/10/2020	17/10/2020
									A072201	High Impedance	17/10/2020	17/10/2020
					CT981	Single use	E2403	No Clinical Signs, Symptoms or Conditions	A070101	Failure to Capture	17/10/2020	17/10/2020
									A070504	Premature Discharge of Battery	17/10/2020	17/10/2020
									A072201	High Impedance	17/10/2020	17/10/2020
							F19	Surgical Intervention	A070101	Failure to Capture	17/10/2020	17/10/2020
									A070504	Premature Discharge of Battery	17/10/2020	17/10/2020
									A072201	High Impedance	17/10/2020	17/10/2020
					CT983	Surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A070101	Failure to Capture	17/10/2020	17/10/2020
									A070504	Premature Discharge of Battery	17/10/2020	17/10/2020
									A072201	High Impedance	17/10/2020	17/10/2020
							F19	Surgical Intervention	A070101	Failure to Capture	17/10/2020	17/10/2020
									A070504	Premature Discharge of Battery	17/10/2020	17/10/2020
									A072201	High Impedance	17/10/2020	17/10/2020

10133	2020	All other reportable incidents	60910	Coronary venous pacing lead	CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A072201	High Impedance		
							F26	No Health Consequences or Impact	A072201	High Impedance		
10134	2019	All other reportable incidents	47263	Cardiac resynchronization therapy implantable pacemaker	CT1684	Cardiac pacemakers and associated devices	F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	07/11/2019	
					CT1685	Cardiac pacemakers	F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	07/11/2019	
					CT257	Implantable cardiac pacemakers	F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	07/11/2019	
					CT2866	Implantable cardiac electrical stimulators	F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	07/11/2019	
10135	2019	All other reportable incidents	47267	Single-chamber implantable pacemaker, rate-responsive	CT141	Electrodes/Leads and associated devices	F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	24/10/2019	24/10/2019
					CT145	Cardiovascular devices	F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	24/10/2019	24/10/2019
					CT1684	Cardiac pacemakers and associated devices	F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	24/10/2019	24/10/2019
					CT1685	Cardiac pacemakers	F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	24/10/2019	24/10/2019
					CT233	Surgical	F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	24/10/2019	24/10/2019
					CT2406	Implantable	F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	24/10/2019	24/10/2019
					CT2407	Active implantable	F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	24/10/2019	24/10/2019
					CT257	Implantable cardiac pacemakers	F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	24/10/2019	24/10/2019
					CT2727	Active implant devices/ancillaries	F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	24/10/2019	24/10/2019
					CT2750	Implantable cardiac electrical stimulators and associated devices	F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	24/10/2019	24/10/2019
					CT2866	Implantable cardiac electrical stimulators	F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	24/10/2019	24/10/2019
					CT2985	Active	F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	24/10/2019	24/10/2019
					CT3001	Implantable metal	F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	24/10/2019	24/10/2019
					CT3002	Implantable synthetic polymer	F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	24/10/2019	24/10/2019
					CT3006	Implantable inorganic material	F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	24/10/2019	24/10/2019
					CT314	Electrically powered	F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	24/10/2019	24/10/2019
					CT321	Long-term surgical invasive	F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	24/10/2019	24/10/2019
					CT449	Generators and associated devices	F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	24/10/2019	24/10/2019
					CT981	Single use	F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	24/10/2019	24/10/2019
					CT983	Surgical invasive	F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	24/10/2019	24/10/2019
10136	2019	All other reportable incidents	36241	Pacemaker/defibrillator lead	CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A072201	High Impedance		
10137	2019	All other reportable incidents	35852	Single-chamber implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	13/11/2019	13/11/2019
10138	2020	All other reportable incidents	35223	Endocardial/interventricular septal pacing lead	CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A0401	Break		
									A07	Electrical /Electronic Property Problem		
									A070909	Oversensing		
									A070910	Undersensing		
									A072202	Low Impedance		
									A090801	Signal Artifact/Noise		
							F26	No Health Consequences or Impact	A0401	Break		
									A07	Electrical /Electronic Property Problem		
									A070909	Oversensing		
									A070910	Undersensing		
									A072202	Low Impedance		
									A090801	Signal Artifact/Noise		
10139	2020	All other reportable incidents	36241	Pacemaker/defibrillator lead	CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A051201	Device Dislodged or Dislocated	24/09/2020	24/09/2020
									A070102	High Capture Threshold	24/09/2020	24/09/2020
							F08	Hospitalization or Prolonged Hospitalization	A051201	Device Dislodged or Dislocated	24/09/2020	24/09/2020
									A070102	High Capture Threshold	24/09/2020	24/09/2020
							F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A051201	Device Dislodged or Dislocated	24/09/2020	24/09/2020
									A070102	High Capture Threshold	24/09/2020	24/09/2020
							F12	Serious Injury/ Illness/ Impairment	A051201	Device Dislodged or Dislocated	24/09/2020	24/09/2020
									A070102	High Capture Threshold	24/09/2020	24/09/2020
							F1903	Device Explantation	A051201	Device Dislodged or Dislocated	24/09/2020	24/09/2020
									A070102	High Capture Threshold	24/09/2020	24/09/2020
							F1905	Device Revision or Replacement	A051201	Device Dislodged or Dislocated	24/09/2020	24/09/2020
									A070102	High Capture Threshold	24/09/2020	24/09/2020
10140	2020	All other reportable incidents	35852	Single-chamber implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A070909	Oversensing		
							F26	No Health Consequences or Impact	A070909	Oversensing		
10141	2020	All other reportable incidents	47270	Cardiac resynchronization therapy implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E2104	Electric Shock	A071302	Inappropriate/Inadequate Shock/Stimulation		
							F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A071302	Inappropriate/Inadequate Shock/Stimulation		
							F2303	Medication Required	A071302	Inappropriate/Inadequate Shock/Stimulation		
10142	2020	All other reportable incidents	35852	Single-chamber implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	06/01/2021	06/01/2021
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	06/01/2021	06/01/2021
							F12	Serious Injury/ Illness/ Impairment	A070504	Premature Discharge of Battery	06/01/2021	06/01/2021
							F19	Surgical Intervention	A070504	Premature Discharge of Battery	06/01/2021	06/01/2021
10143	2020	All other reportable incidents	47265	Dual-chamber implantable pacemaker, rate-responsive	CT141	Electrodes/Leads and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	22/06/2021	22/06/2021
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	22/06/2021	22/06/2021
					CT145	Cardiovascular devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	22/06/2021	22/06/2021
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	22/06/2021	22/06/2021
					CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	22/06/2021	22/06/2021
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	22/06/2021	22/06/2021
					CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	22/06/2021	22/06/2021
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	22/06/2021	22/06/2021
					CT233	Surgical	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	22/06/2021	22/06/2021
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	22/06/2021	22/06/2021
					CT2406	Implantable	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	22/06/2021	22/06/2021
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	22/06/2021	22/06/2021
					CT2407	Active implantable	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	22/06/2021	22/06/2021
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	22/06/2021	22/06/2021
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	22/06/2021	22/06/2021
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	22/06/2021	22/06/2021
					CT2727	Active implant devices/ancillaries	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	22/06/2021	22/06/2021
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	22/06/2021	22/06/2021
					CT2750	Implantable cardiac electrical stimulators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	22/06/2021	22/06/2021
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	22/06/2021	22/06/2021
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	22/06/2021	22/06/2021
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	22/06/2021	22/06/2021
					CT2985	Active	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	22/06/2021	22/06/2021
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	22/06/2021	22/06/2021
					CT3001	Implantable metal	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	22/06/2021	22/06/2021
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	22/06/2021	22/06/2021

					CT3002	Implantable synthetic polymer	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	22/06/2021	22/06/2021
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	22/06/2021	22/06/2021
					CT3006	Implantable inorganic material	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	22/06/2021	22/06/2021
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	22/06/2021	22/06/2021
					CT314	Electrically powered	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	22/06/2021	22/06/2021
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	22/06/2021	22/06/2021
					CT321	Long-term surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	22/06/2021	22/06/2021
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	22/06/2021	22/06/2021
					CT449	Generators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	22/06/2021	22/06/2021
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	22/06/2021	22/06/2021
					CT981	Single use	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	22/06/2021	22/06/2021
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	22/06/2021	22/06/2021
					CT983	Surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	22/06/2021	22/06/2021
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	22/06/2021	22/06/2021
10144	2020	All other reportable incidents	47265	Dual-chamber implantable pacemaker, rate-responsive	CT141	Electrodes/Leads and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT145	Cardiovascular devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT233	Surgical	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT2406	Implantable	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT2407	Active implantable	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT2727	Active implant devices/ancillaries	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT2750	Implantable cardiac electrical stimulators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT2985	Active	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT3001	Implantable metal	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT3002	Implantable synthetic polymer	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT3006	Implantable inorganic material	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT314	Electrically powered	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT321	Long-term surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT449	Generators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT981	Single use	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT983	Surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
10145	2019	All other reportable incidents	35852	Single-chamber implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	F1905	Device Revision or Replacement	A0705	Battery Problem	24/10/2019	
10146	2020	All other reportable incidents	47265	Dual-chamber implantable pacemaker, rate-responsive	CT141	Electrodes/Leads and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	01/10/2020	01/10/2020
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	01/10/2020	01/10/2020
							F12	Serious Injury/ Illness/ Impairment	A070504	Premature Discharge of Battery	01/10/2020	01/10/2020
							F1903	Device Explantation	A070504	Premature Discharge of Battery	01/10/2020	01/10/2020
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	01/10/2020	01/10/2020
					CT145	Cardiovascular devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	01/10/2020	01/10/2020
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	01/10/2020	01/10/2020
							F12	Serious Injury/ Illness/ Impairment	A070504	Premature Discharge of Battery	01/10/2020	01/10/2020
							F1903	Device Explantation	A070504	Premature Discharge of Battery	01/10/2020	01/10/2020
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	01/10/2020	01/10/2020
					CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	01/10/2020	01/10/2020
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	01/10/2020	01/10/2020
							F12	Serious Injury/ Illness/ Impairment	A070504	Premature Discharge of Battery	01/10/2020	01/10/2020
							F1903	Device Explantation	A070504	Premature Discharge of Battery	01/10/2020	01/10/2020
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	01/10/2020	01/10/2020
					CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	01/10/2020	01/10/2020
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	01/10/2020	01/10/2020
							F12	Serious Injury/ Illness/ Impairment	A070504	Premature Discharge of Battery	01/10/2020	01/10/2020
							F1903	Device Explantation	A070504	Premature Discharge of Battery	01/10/2020	01/10/2020
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	01/10/2020	01/10/2020
					CT233	Surgical	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	01/10/2020	01/10/2020
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	01/10/2020	01/10/2020
							F12	Serious Injury/ Illness/ Impairment	A070504	Premature Discharge of Battery	01/10/2020	01/10/2020
							F1903	Device Explantation	A070504	Premature Discharge of Battery	01/10/2020	01/10/2020
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	01/10/2020	01/10/2020
					CT2406	Implantable	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	01/10/2020	01/10/2020
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	01/10/2020	01/10/2020
							F12	Serious Injury/ Illness/ Impairment	A070504	Premature Discharge of Battery	01/10/2020	01/10/2020
							F1903	Device Explantation	A070504	Premature Discharge of Battery	01/10/2020	01/10/2020
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	01/10/2020	01/10/2020
					CT2407	Active implantable	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	01/10/2020	01/10/2020
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	01/10/2020	01/10/2020
							F12	Serious Injury/ Illness/ Impairment	A070504	Premature Discharge of Battery	01/10/2020	01/10/2020

							F1903	Device Explantation	A070504	Premature Discharge of Battery	01/10/2020	01/10/2020
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	01/10/2020	01/10/2020
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	01/10/2020	01/10/2020
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	01/10/2020	01/10/2020
							F12	Serious Injury/ Illness/ Impairment	A070504	Premature Discharge of Battery	01/10/2020	01/10/2020
							F1903	Device Explantation	A070504	Premature Discharge of Battery	01/10/2020	01/10/2020
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	01/10/2020	01/10/2020
					CT2727	Active implant devices/ancillaries	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	01/10/2020	01/10/2020
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	01/10/2020	01/10/2020
							F12	Serious Injury/ Illness/ Impairment	A070504	Premature Discharge of Battery	01/10/2020	01/10/2020
							F1903	Device Explantation	A070504	Premature Discharge of Battery	01/10/2020	01/10/2020
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	01/10/2020	01/10/2020
					CT2750	Implantable cardiac electrical stimulators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	01/10/2020	01/10/2020
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	01/10/2020	01/10/2020
							F12	Serious Injury/ Illness/ Impairment	A070504	Premature Discharge of Battery	01/10/2020	01/10/2020
							F1903	Device Explantation	A070504	Premature Discharge of Battery	01/10/2020	01/10/2020
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	01/10/2020	01/10/2020
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	01/10/2020	01/10/2020
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	01/10/2020	01/10/2020
							F12	Serious Injury/ Illness/ Impairment	A070504	Premature Discharge of Battery	01/10/2020	01/10/2020
							F1903	Device Explantation	A070504	Premature Discharge of Battery	01/10/2020	01/10/2020
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	01/10/2020	01/10/2020
					CT2985	Active	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	01/10/2020	01/10/2020
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	01/10/2020	01/10/2020
							F12	Serious Injury/ Illness/ Impairment	A070504	Premature Discharge of Battery	01/10/2020	01/10/2020
							F1903	Device Explantation	A070504	Premature Discharge of Battery	01/10/2020	01/10/2020
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	01/10/2020	01/10/2020
					CT3001	Implantable metal	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	01/10/2020	01/10/2020
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	01/10/2020	01/10/2020
							F12	Serious Injury/ Illness/ Impairment	A070504	Premature Discharge of Battery	01/10/2020	01/10/2020
							F1903	Device Explantation	A070504	Premature Discharge of Battery	01/10/2020	01/10/2020
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	01/10/2020	01/10/2020
					CT3002	Implantable synthetic polymer	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	01/10/2020	01/10/2020
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	01/10/2020	01/10/2020
							F12	Serious Injury/ Illness/ Impairment	A070504	Premature Discharge of Battery	01/10/2020	01/10/2020
							F1903	Device Explantation	A070504	Premature Discharge of Battery	01/10/2020	01/10/2020
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	01/10/2020	01/10/2020
					CT3006	Implantable inorganic material	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	01/10/2020	01/10/2020
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	01/10/2020	01/10/2020
							F12	Serious Injury/ Illness/ Impairment	A070504	Premature Discharge of Battery	01/10/2020	01/10/2020
							F1903	Device Explantation	A070504	Premature Discharge of Battery	01/10/2020	01/10/2020
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	01/10/2020	01/10/2020
					CT314	Electrically powered	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	01/10/2020	01/10/2020
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	01/10/2020	01/10/2020
							F12	Serious Injury/ Illness/ Impairment	A070504	Premature Discharge of Battery	01/10/2020	01/10/2020
							F1903	Device Explantation	A070504	Premature Discharge of Battery	01/10/2020	01/10/2020
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	01/10/2020	01/10/2020
					CT321	Long-term surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	01/10/2020	01/10/2020
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	01/10/2020	01/10/2020
							F12	Serious Injury/ Illness/ Impairment	A070504	Premature Discharge of Battery	01/10/2020	01/10/2020
							F1903	Device Explantation	A070504	Premature Discharge of Battery	01/10/2020	01/10/2020
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	01/10/2020	01/10/2020
					CT449	Generators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	01/10/2020	01/10/2020
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	01/10/2020	01/10/2020
							F12	Serious Injury/ Illness/ Impairment	A070504	Premature Discharge of Battery	01/10/2020	01/10/2020
							F1903	Device Explantation	A070504	Premature Discharge of Battery	01/10/2020	01/10/2020
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	01/10/2020	01/10/2020
					CT981	Single use	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	01/10/2020	01/10/2020
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	01/10/2020	01/10/2020
							F12	Serious Injury/ Illness/ Impairment	A070504	Premature Discharge of Battery	01/10/2020	01/10/2020
							F1903	Device Explantation	A070504	Premature Discharge of Battery	01/10/2020	01/10/2020
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	01/10/2020	01/10/2020
					CT983	Surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	01/10/2020	01/10/2020
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	01/10/2020	01/10/2020
							F12	Serious Injury/ Illness/ Impairment	A070504	Premature Discharge of Battery	01/10/2020	01/10/2020
							F1903	Device Explantation	A070504	Premature Discharge of Battery	01/10/2020	01/10/2020
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	01/10/2020	01/10/2020
10147	2020	All other reportable incidents	35852	Single-chamber implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E060102	Atrial Fibrillation	A072201	High Impedance		
							E2403	No Clinical Signs, Symptoms or Conditions	A072201	High Impedance		
							F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A072201	High Impedance		
							F26	No Health Consequences or Impact	A072201	High Impedance		
10148	2020	All other reportable incidents	37265	Dual-chamber implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E060102	Atrial Fibrillation	A0712	Pacing Problem		
									A071302	Inappropriate/Inadequate Shock/Stimulation		
									A0909	Unexpected Therapeutic Results		
							E2104	Electric Shock	A0712	Pacing Problem		
									A071302	Inappropriate/Inadequate Shock/Stimulation		
									A0909	Unexpected Therapeutic Results		
							F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A0712	Pacing Problem		
									A071302	Inappropriate/Inadequate Shock/Stimulation		
									A0909	Unexpected Therapeutic Results		
							F26	No Health Consequences or Impact	A0712	Pacing Problem		
									A071302	Inappropriate/Inadequate Shock/Stimulation		
									A0909	Unexpected Therapeutic Results		
10149	2020	All other reportable incidents	47265	Dual-chamber implantable pacemaker, rate-responsive	CT141	Electrodes/Leads and associated devices	E060103	Atrial Flutter	A070910	Undersensing		
									A0712	Pacing Problem		
							E0623	Unspecified Heart Problem	A070910	Undersensing		
									A0712	Pacing Problem		
							F26	No Health Consequences or Impact	A070910	Undersensing		

								A0712	Pacing Problem		
					CT145	Cardiovascular devices	E060103	Atrial Flutter	A070910	Undersensing		
									A0712	Pacing Problem		
							E0623	Unspecified Heart Problem	A070910	Undersensing		
									A0712	Pacing Problem		
							F26	No Health Consequences or Impact	A070910	Undersensing		
									A0712	Pacing Problem		
					CT1684	Cardiac pacemakers and associated devices	E060103	Atrial Flutter	A070910	Undersensing		
									A0712	Pacing Problem		
							E0623	Unspecified Heart Problem	A070910	Undersensing		
									A0712	Pacing Problem		
							F26	No Health Consequences or Impact	A070910	Undersensing		
									A0712	Pacing Problem		
					CT1685	Cardiac pacemakers	E060103	Atrial Flutter	A070910	Undersensing		
									A0712	Pacing Problem		
							E0623	Unspecified Heart Problem	A070910	Undersensing		
									A0712	Pacing Problem		
							F26	No Health Consequences or Impact	A070910	Undersensing		
									A0712	Pacing Problem		
					CT233	Surgical	E060103	Atrial Flutter	A070910	Undersensing		
									A0712	Pacing Problem		
							E0623	Unspecified Heart Problem	A070910	Undersensing		
									A0712	Pacing Problem		
							F26	No Health Consequences or Impact	A070910	Undersensing		
									A0712	Pacing Problem		
					CT2406	Implantable	E060103	Atrial Flutter	A070910	Undersensing		
									A0712	Pacing Problem		
							E0623	Unspecified Heart Problem	A070910	Undersensing		
									A0712	Pacing Problem		
							F26	No Health Consequences or Impact	A070910	Undersensing		
									A0712	Pacing Problem		
					CT2407	Active implantable	E060103	Atrial Flutter	A070910	Undersensing		
									A0712	Pacing Problem		
							E0623	Unspecified Heart Problem	A070910	Undersensing		
									A0712	Pacing Problem		
							F26	No Health Consequences or Impact	A070910	Undersensing		
									A0712	Pacing Problem		
					CT257	Implantable cardiac pacemakers	E060103	Atrial Flutter	A070910	Undersensing		
									A0712	Pacing Problem		
							E0623	Unspecified Heart Problem	A070910	Undersensing		
									A0712	Pacing Problem		
							F26	No Health Consequences or Impact	A070910	Undersensing		
									A0712	Pacing Problem		
					CT2727	Active implant devices/ancillaries	E060103	Atrial Flutter	A070910	Undersensing		
									A0712	Pacing Problem		
							E0623	Unspecified Heart Problem	A070910	Undersensing		
									A0712	Pacing Problem		
							F26	No Health Consequences or Impact	A070910	Undersensing		
									A0712	Pacing Problem		
					CT2750	Implantable cardiac electrical stimulators and associated devices	E060103	Atrial Flutter	A070910	Undersensing		
									A0712	Pacing Problem		
							E0623	Unspecified Heart Problem	A070910	Undersensing		
									A0712	Pacing Problem		
							F26	No Health Consequences or Impact	A070910	Undersensing		
									A0712	Pacing Problem		
					CT2866	Implantable cardiac electrical stimulators	E060103	Atrial Flutter	A070910	Undersensing		
									A0712	Pacing Problem		
							E0623	Unspecified Heart Problem	A070910	Undersensing		
									A0712	Pacing Problem		
							F26	No Health Consequences or Impact	A070910	Undersensing		
									A0712	Pacing Problem		
					CT2985	Active	E060103	Atrial Flutter	A070910	Undersensing		
									A0712	Pacing Problem		
							E0623	Unspecified Heart Problem	A070910	Undersensing		
									A0712	Pacing Problem		
							F26	No Health Consequences or Impact	A070910	Undersensing		
									A0712	Pacing Problem		
					CT3001	Implantable metal	E060103	Atrial Flutter	A070910	Undersensing		
									A0712	Pacing Problem		
							E0623	Unspecified Heart Problem	A070910	Undersensing		
									A0712	Pacing Problem		
							F26	No Health Consequences or Impact	A070910	Undersensing		
									A0712	Pacing Problem		
					CT3002	Implantable synthetic polymer	E060103	Atrial Flutter	A070910	Undersensing		
									A0712	Pacing Problem		
							E0623	Unspecified Heart Problem	A070910	Undersensing		
									A0712	Pacing Problem		
							F26	No Health Consequences or Impact	A070910	Undersensing		
									A0712	Pacing Problem		
					CT3006	Implantable inorganic material	E060103	Atrial Flutter	A070910	Undersensing		
									A0712	Pacing Problem		
							E0623	Unspecified Heart Problem	A070910	Undersensing		
									A0712	Pacing Problem		
							F26	No Health Consequences or Impact	A070910	Undersensing		
									A0712	Pacing Problem		
					CT314	Electrically powered	E060103	Atrial Flutter	A070910	Undersensing		
									A0712	Pacing Problem		
							E0623	Unspecified Heart Problem	A070910	Undersensing		

									A0712	Pacing Problem		
							F26	No Health Consequences or Impact	A070910	Undersensing		
									A0712	Pacing Problem		
					CT321	Long-term surgical invasive	E060103	Atrial Flutter	A070910	Undersensing		
									A0712	Pacing Problem		
							E0623	Unspecified Heart Problem	A070910	Undersensing		
									A0712	Pacing Problem		
							F26	No Health Consequences or Impact	A070910	Undersensing		
									A0712	Pacing Problem		
					CT449	Generators and associated devices	E060103	Atrial Flutter	A070910	Undersensing		
									A0712	Pacing Problem		
							E0623	Unspecified Heart Problem	A070910	Undersensing		
									A0712	Pacing Problem		
							F26	No Health Consequences or Impact	A070910	Undersensing		
									A0712	Pacing Problem		
					CT981	Single use	E060103	Atrial Flutter	A070910	Undersensing		
									A0712	Pacing Problem		
							E0623	Unspecified Heart Problem	A070910	Undersensing		
									A0712	Pacing Problem		
							F26	No Health Consequences or Impact	A070910	Undersensing		
									A0712	Pacing Problem		
					CT983	Surgical invasive	E060103	Atrial Flutter	A070910	Undersensing		
									A0712	Pacing Problem		
							E0623	Unspecified Heart Problem	A070910	Undersensing		
									A0712	Pacing Problem		
							F26	No Health Consequences or Impact	A070910	Undersensing		
									A0712	Pacing Problem		
10150	2020	All other reportable incidents	47265	Dual-chamber implantable pacemaker, rate-responsive	CT141	Electrodes/Leads and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT145	Cardiovascular devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT233	Surgical	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT2406	Implantable	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT2407	Active implantable	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT2727	Active implant devices/ancillaries	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT2750	Implantable cardiac electrical stimulators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT2985	Active	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT3001	Implantable metal	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT3002	Implantable synthetic polymer	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT3006	Implantable inorganic material	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT314	Electrically powered	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT321	Long-term surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT449	Generators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT981	Single use	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT983	Surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
10151	2020	All other reportable incidents	37265	Dual-chamber implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
10152	2019	All other reportable incidents	35852	Single-chamber implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A072201	High Impedance		
10153	2019	All other reportable incidents	36241	Pacemaker/defibrillator lead	CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A072201	High Impedance		
10154	2019	All other reportable incidents	37265	Dual-chamber implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A070909	Oversensing		
10155	2019	All other reportable incidents	37265	Dual-chamber implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A071302	Inappropriate/Inadequate Shock/Stimulation		
10156	2020	All other reportable incidents	47265	Dual-chamber implantable pacemaker, rate-responsive	CT141	Electrodes/Leads and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	05/10/2020	21/01/2021
							F19	Surgical Intervention	A070504	Premature Discharge of Battery	05/10/2020	21/01/2021
					CT145	Cardiovascular devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	05/10/2020	21/01/2021
							F19	Surgical Intervention	A070504	Premature Discharge of Battery	05/10/2020	21/01/2021
					CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	05/10/2020	21/01/2021
							F19	Surgical Intervention	A070504	Premature Discharge of Battery	05/10/2020	21/01/2021
					CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	05/10/2020	21/01/2021
							F19	Surgical Intervention	A070504	Premature Discharge of Battery	05/10/2020	21/01/2021
					CT233	Surgical	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	05/10/2020	21/01/2021
							F19	Surgical Intervention	A070504	Premature Discharge of Battery	05/10/2020	21/01/2021
					CT2406	Implantable	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	05/10/2020	21/01/2021
							F19	Surgical Intervention	A070504	Premature Discharge of Battery	05/10/2020	21/01/2021
					CT2407	Active implantable	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	05/10/2020	21/01/2021
							F19	Surgical Intervention	A070504	Premature Discharge of Battery	05/10/2020	21/01/2021

					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	05/10/2020	21/01/2021
							F19	Surgical Intervention	A070504	Premature Discharge of Battery	05/10/2020	21/01/2021
					CT2727	Active implant devices/ancillaries	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	05/10/2020	21/01/2021
							F19	Surgical Intervention	A070504	Premature Discharge of Battery	05/10/2020	21/01/2021
					CT2750	Implantable cardiac electrical stimulators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	05/10/2020	21/01/2021
							F19	Surgical Intervention	A070504	Premature Discharge of Battery	05/10/2020	21/01/2021
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	05/10/2020	21/01/2021
							F19	Surgical Intervention	A070504	Premature Discharge of Battery	05/10/2020	21/01/2021
					CT2985	Active	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	05/10/2020	21/01/2021
							F19	Surgical Intervention	A070504	Premature Discharge of Battery	05/10/2020	21/01/2021
					CT3001	Implantable metal	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	05/10/2020	21/01/2021
							F19	Surgical Intervention	A070504	Premature Discharge of Battery	05/10/2020	21/01/2021
					CT3002	Implantable synthetic polymer	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	05/10/2020	21/01/2021
							F19	Surgical Intervention	A070504	Premature Discharge of Battery	05/10/2020	21/01/2021
					CT3006	Implantable inorganic material	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	05/10/2020	21/01/2021
							F19	Surgical Intervention	A070504	Premature Discharge of Battery	05/10/2020	21/01/2021
					CT314	Electrically powered	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	05/10/2020	21/01/2021
							F19	Surgical Intervention	A070504	Premature Discharge of Battery	05/10/2020	21/01/2021
					CT321	Long-term surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	05/10/2020	21/01/2021
							F19	Surgical Intervention	A070504	Premature Discharge of Battery	05/10/2020	21/01/2021
					CT449	Generators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	05/10/2020	21/01/2021
							F19	Surgical Intervention	A070504	Premature Discharge of Battery	05/10/2020	21/01/2021
					CT981	Single use	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	05/10/2020	21/01/2021
							F19	Surgical Intervention	A070504	Premature Discharge of Battery	05/10/2020	21/01/2021
					CT983	Surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	05/10/2020	21/01/2021
							F19	Surgical Intervention	A070504	Premature Discharge of Battery	05/10/2020	21/01/2021
10157	2020	All other reportable incidents	47270	Cardiac resynchronization therapy implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A070909	Oversensing		
									A0712	Pacing Problem		
									A090801	Signal Artifact/Noise		
							F26	No Health Consequences or Impact	A070909	Oversensing		
									A0712	Pacing Problem		
									A090801	Signal Artifact/Noise		
10158	2020	All other reportable incidents	35852	Single-chamber implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A072201	High Impedance		
							F26	No Health Consequences or Impact	A072201	High Impedance		
10159	2020	All other reportable incidents	36241	Pacemaker/defibrillator lead	CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A0203	Defective Device	15/10/2020	15/10/2020
									A070909	Oversensing	15/10/2020	15/10/2020
									A0712	Pacing Problem	15/10/2020	15/10/2020
									A072202	Low Impedance	15/10/2020	15/10/2020
									A090801	Signal Artifact/Noise	15/10/2020	15/10/2020
							F08	Hospitalization or Prolonged Hospitalization	A0203	Defective Device	15/10/2020	15/10/2020
									A070909	Oversensing	15/10/2020	15/10/2020
									A0712	Pacing Problem	15/10/2020	15/10/2020
									A072202	Low Impedance	15/10/2020	15/10/2020
									A090801	Signal Artifact/Noise	15/10/2020	15/10/2020
							F19	Surgical Intervention	A0203	Defective Device	15/10/2020	15/10/2020
									A070909	Oversensing	15/10/2020	15/10/2020
									A0712	Pacing Problem	15/10/2020	15/10/2020
									A072202	Low Impedance	15/10/2020	15/10/2020
									A090801	Signal Artifact/Noise	15/10/2020	15/10/2020
							F1905	Device Revision or Replacement	A0203	Defective Device	15/10/2020	15/10/2020
									A070909	Oversensing	15/10/2020	15/10/2020
									A0712	Pacing Problem	15/10/2020	15/10/2020
									A072202	Low Impedance	15/10/2020	15/10/2020
									A090801	Signal Artifact/Noise	15/10/2020	15/10/2020
10160	2020	All other reportable incidents	36241	Pacemaker/defibrillator lead	CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A040101	Fracture		
									A0712	Pacing Problem		
									A072201	High Impedance		
									A090801	Signal Artifact/Noise		
							F08	Hospitalization or Prolonged Hospitalization	A040101	Fracture		
									A0712	Pacing Problem		
									A072201	High Impedance		
									A090801	Signal Artifact/Noise		
							F12	Serious Injury/ Illness/ Impairment	A040101	Fracture		
									A0712	Pacing Problem		
									A072201	High Impedance		
									A090801	Signal Artifact/Noise		
							F19	Surgical Intervention	A040101	Fracture		
									A0712	Pacing Problem		
									A072201	High Impedance		
									A090801	Signal Artifact/Noise		
							F1905	Device Revision or Replacement	A040101	Fracture		
									A0712	Pacing Problem		
									A072201	High Impedance		
									A090801	Signal Artifact/Noise		
10161	2020	All other reportable incidents	17882	Physiologic-monitoring defibrillation system	CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070803	Failure to Power Up		
							F27	Problem Identified During Non-Clinical Procedure	A070803	Failure to Power Up		
10162	2019	All other reportable incidents	37265	Dual-chamber implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	07/11/2019	
10163	2019	All other reportable incidents	47270	Cardiac resynchronization therapy implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A110206	Application Program Version or Upgrade Problem		
10164	2019	All other reportable incidents	47270	Cardiac resynchronization therapy implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	05/11/2019	
10165	2019	All other reportable incidents	37265	Dual-chamber implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	12/11/2019	
10166	2019	All other reportable incidents	35852	Single-chamber implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	23/09/2019	
10167	2019	All other reportable incidents	47265	Dual-chamber implantable pacemaker, rate-responsive	CT141	Electrodes/Leads and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A072201	High Impedance		
							F26	No Health Consequences or Impact	A072201	High Impedance		
					CT145	Cardiovascular devices	E2403	No Clinical Signs, Symptoms or Conditions	A072201	High Impedance		
							F26	No Health Consequences or Impact	A072201	High Impedance		
					CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A072201	High Impedance		

							F26	No Health Consequences or Impact	A072201	High Impedance			
					CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A072201	High Impedance			
							F26	No Health Consequences or Impact	A072201	High Impedance			
					CT233	Surgical	E2403	No Clinical Signs, Symptoms or Conditions	A072201	High Impedance			
							F26	No Health Consequences or Impact	A072201	High Impedance			
					CT2406	Implantable	E2403	No Clinical Signs, Symptoms or Conditions	A072201	High Impedance			
							F26	No Health Consequences or Impact	A072201	High Impedance			
					CT2407	Active implantable	E2403	No Clinical Signs, Symptoms or Conditions	A072201	High Impedance			
							F26	No Health Consequences or Impact	A072201	High Impedance			
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A072201	High Impedance			
							F26	No Health Consequences or Impact	A072201	High Impedance			
					CT2727	Active implant devices/ancillaries	E2403	No Clinical Signs, Symptoms or Conditions	A072201	High Impedance			
							F26	No Health Consequences or Impact	A072201	High Impedance			
					CT2750	Implantable cardiac electrical stimulators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A072201	High Impedance			
							F26	No Health Consequences or Impact	A072201	High Impedance			
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A072201	High Impedance			
							F26	No Health Consequences or Impact	A072201	High Impedance			
					CT2985	Active	E2403	No Clinical Signs, Symptoms or Conditions	A072201	High Impedance			
							F26	No Health Consequences or Impact	A072201	High Impedance			
					CT3001	Implantable metal	E2403	No Clinical Signs, Symptoms or Conditions	A072201	High Impedance			
							F26	No Health Consequences or Impact	A072201	High Impedance			
					CT3002	Implantable synthetic polymer	E2403	No Clinical Signs, Symptoms or Conditions	A072201	High Impedance			
							F26	No Health Consequences or Impact	A072201	High Impedance			
					CT3006	Implantable inorganic material	E2403	No Clinical Signs, Symptoms or Conditions	A072201	High Impedance			
							F26	No Health Consequences or Impact	A072201	High Impedance			
					CT314	Electrically powered	E2403	No Clinical Signs, Symptoms or Conditions	A072201	High Impedance			
							F26	No Health Consequences or Impact	A072201	High Impedance			
					CT321	Long-term surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A072201	High Impedance			
							F26	No Health Consequences or Impact	A072201	High Impedance			
					CT449	Generators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A072201	High Impedance			
							F26	No Health Consequences or Impact	A072201	High Impedance			
					CT981	Single use	E2403	No Clinical Signs, Symptoms or Conditions	A072201	High Impedance			
							F26	No Health Consequences or Impact	A072201	High Impedance			
					CT983	Surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A072201	High Impedance			
							F26	No Health Consequences or Impact	A072201	High Impedance			
10168	2019	All other reportable incidents	37265	Dual-chamber implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A070909	Oversensing			
10169	2019	All other reportable incidents	47265	Dual-chamber implantable pacemaker, rate-responsive	CT141	Electrodes/Leads and associated devices	F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	23/01/2020	23/01/2020	
					CT145	Cardiovascular devices	F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	23/01/2020	23/01/2020	
					CT1684	Cardiac pacemakers and associated devices	F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	23/01/2020	23/01/2020	
					CT1685	Cardiac pacemakers	F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	23/01/2020	23/01/2020	
					CT233	Surgical	F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	23/01/2020	23/01/2020	
					CT2406	Implantable	F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	23/01/2020	23/01/2020	
					CT2407	Active implantable	F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	23/01/2020	23/01/2020	
					CT257	Implantable cardiac pacemakers	F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	23/01/2020	23/01/2020	
					CT2727	Active implant devices/ancillaries	F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	23/01/2020	23/01/2020	
					CT2750	Implantable cardiac electrical stimulators and associated devices	F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	23/01/2020	23/01/2020	
					CT2866	Implantable cardiac electrical stimulators	F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	23/01/2020	23/01/2020	
					CT2985	Active	F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	23/01/2020	23/01/2020	
					CT3001	Implantable metal	F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	23/01/2020	23/01/2020	
					CT3002	Implantable synthetic polymer	F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	23/01/2020	23/01/2020	
					CT3006	Implantable inorganic material	F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	23/01/2020	23/01/2020	
					CT314	Electrically powered	F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	23/01/2020	23/01/2020	
					CT321	Long-term surgical invasive	F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	23/01/2020	23/01/2020	
					CT449	Generators and associated devices	F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	23/01/2020	23/01/2020	
					CT981	Single use	F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	23/01/2020	23/01/2020	
					CT983	Surgical invasive	F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	23/01/2020	23/01/2020	
10170	2019	All other reportable incidents	35852	Single-chamber implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A070909	Oversensing			
10171	2019	All other reportable incidents	35852	Single-chamber implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A070909	Oversensing			
10172	2020	All other reportable incidents	47270	Cardiac resynchronization therapy implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E2104	Electric Shock	A071302	Inappropriate/Inadequate Shock/Stimulation			
									A090801	Signal Artifact/Noise			
							F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A071302	Inappropriate/Inadequate Shock/Stimulation			
									A090801	Signal Artifact/Noise			
							F15	Recognized Device or Procedural Complication	A071302	Inappropriate/Inadequate Shock/Stimulation			
									A090801	Signal Artifact/Noise			
10173	2020	All other reportable incidents	36241	Pacemaker/defibrillator lead	CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070102	High Capture Threshold			
									A070910	Undersensing			
									A090801	Signal Artifact/Noise			
							F15	Recognized Device or Procedural Complication	A070102	High Capture Threshold			
									A070910	Undersensing			
									A090801	Signal Artifact/Noise			
10174	2020	All other reportable incidents	47265	Dual-chamber implantable pacemaker, rate-responsive	CT141	Electrodes/Leads and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	30/09/2020	30/09/2020	
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	30/09/2020	30/09/2020	
							F12	Serious Injury/ Illness/ Impairment	A070504	Premature Discharge of Battery	30/09/2020	30/09/2020	
							F19	Surgical Intervention	A070504	Premature Discharge of Battery	30/09/2020	30/09/2020	
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	30/09/2020	30/09/2020	
					CT145	Cardiovascular devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	30/09/2020	30/09/2020	
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	30/09/2020	30/09/2020	
							F12	Serious Injury/ Illness/ Impairment	A070504	Premature Discharge of Battery	30/09/2020	30/09/2020	
							F19	Surgical Intervention	A070504	Premature Discharge of Battery	30/09/2020	30/09/2020	
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	30/09/2020	30/09/2020	
					CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	30/09/2020	30/09/2020	
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	30/09/2020	30/09/2020	
							F12	Serious Injury/ Illness/ Impairment	A070504	Premature Discharge of Battery	30/09/2020	30/09/2020	
							F19	Surgical Intervention	A070504	Premature Discharge of Battery	30/09/2020	30/09/2020	
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	30/09/2020	30/09/2020	
					CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	30/09/2020	30/09/2020	
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	30/09/2020	30/09/2020	

							F12	Serious Injury/ Illness/ Impairment	A070504	Premature Discharge of Battery	30/09/2020	30/09/2020
							F19	Surgical Intervention	A070504	Premature Discharge of Battery	30/09/2020	30/09/2020
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	30/09/2020	30/09/2020
					CT233	Surgical	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	30/09/2020	30/09/2020
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	30/09/2020	30/09/2020
							F12	Serious Injury/ Illness/ Impairment	A070504	Premature Discharge of Battery	30/09/2020	30/09/2020
							F19	Surgical Intervention	A070504	Premature Discharge of Battery	30/09/2020	30/09/2020
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	30/09/2020	30/09/2020
					CT2406	Implantable	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	30/09/2020	30/09/2020
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	30/09/2020	30/09/2020
							F12	Serious Injury/ Illness/ Impairment	A070504	Premature Discharge of Battery	30/09/2020	30/09/2020
							F19	Surgical Intervention	A070504	Premature Discharge of Battery	30/09/2020	30/09/2020
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	30/09/2020	30/09/2020
					CT2407	Active implantable	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	30/09/2020	30/09/2020
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	30/09/2020	30/09/2020
							F12	Serious Injury/ Illness/ Impairment	A070504	Premature Discharge of Battery	30/09/2020	30/09/2020
							F19	Surgical Intervention	A070504	Premature Discharge of Battery	30/09/2020	30/09/2020
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	30/09/2020	30/09/2020
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	30/09/2020	30/09/2020
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	30/09/2020	30/09/2020
							F12	Serious Injury/ Illness/ Impairment	A070504	Premature Discharge of Battery	30/09/2020	30/09/2020
							F19	Surgical Intervention	A070504	Premature Discharge of Battery	30/09/2020	30/09/2020
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	30/09/2020	30/09/2020
					CT2727	Active implant devices/ancillaries	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	30/09/2020	30/09/2020
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	30/09/2020	30/09/2020
							F12	Serious Injury/ Illness/ Impairment	A070504	Premature Discharge of Battery	30/09/2020	30/09/2020
							F19	Surgical Intervention	A070504	Premature Discharge of Battery	30/09/2020	30/09/2020
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	30/09/2020	30/09/2020
					CT2750	Implantable cardiac electrical stimulators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	30/09/2020	30/09/2020
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	30/09/2020	30/09/2020
							F12	Serious Injury/ Illness/ Impairment	A070504	Premature Discharge of Battery	30/09/2020	30/09/2020
							F19	Surgical Intervention	A070504	Premature Discharge of Battery	30/09/2020	30/09/2020
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	30/09/2020	30/09/2020
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	30/09/2020	30/09/2020
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	30/09/2020	30/09/2020
							F12	Serious Injury/ Illness/ Impairment	A070504	Premature Discharge of Battery	30/09/2020	30/09/2020
							F19	Surgical Intervention	A070504	Premature Discharge of Battery	30/09/2020	30/09/2020
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	30/09/2020	30/09/2020
					CT2985	Active	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	30/09/2020	30/09/2020
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	30/09/2020	30/09/2020
							F12	Serious Injury/ Illness/ Impairment	A070504	Premature Discharge of Battery	30/09/2020	30/09/2020
							F19	Surgical Intervention	A070504	Premature Discharge of Battery	30/09/2020	30/09/2020
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	30/09/2020	30/09/2020
					CT3001	Implantable metal	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	30/09/2020	30/09/2020
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	30/09/2020	30/09/2020
							F12	Serious Injury/ Illness/ Impairment	A070504	Premature Discharge of Battery	30/09/2020	30/09/2020
							F19	Surgical Intervention	A070504	Premature Discharge of Battery	30/09/2020	30/09/2020
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	30/09/2020	30/09/2020
					CT3002	Implantable synthetic polymer	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	30/09/2020	30/09/2020
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	30/09/2020	30/09/2020
							F12	Serious Injury/ Illness/ Impairment	A070504	Premature Discharge of Battery	30/09/2020	30/09/2020
							F19	Surgical Intervention	A070504	Premature Discharge of Battery	30/09/2020	30/09/2020
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	30/09/2020	30/09/2020
					CT3006	Implantable inorganic material	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	30/09/2020	30/09/2020
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	30/09/2020	30/09/2020
							F12	Serious Injury/ Illness/ Impairment	A070504	Premature Discharge of Battery	30/09/2020	30/09/2020
							F19	Surgical Intervention	A070504	Premature Discharge of Battery	30/09/2020	30/09/2020
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	30/09/2020	30/09/2020
					CT314	Electrically powered	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	30/09/2020	30/09/2020
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	30/09/2020	30/09/2020
							F12	Serious Injury/ Illness/ Impairment	A070504	Premature Discharge of Battery	30/09/2020	30/09/2020
							F19	Surgical Intervention	A070504	Premature Discharge of Battery	30/09/2020	30/09/2020
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	30/09/2020	30/09/2020
					CT321	Long-term surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	30/09/2020	30/09/2020
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	30/09/2020	30/09/2020
							F12	Serious Injury/ Illness/ Impairment	A070504	Premature Discharge of Battery	30/09/2020	30/09/2020
							F19	Surgical Intervention	A070504	Premature Discharge of Battery	30/09/2020	30/09/2020
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	30/09/2020	30/09/2020
					CT449	Generators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	30/09/2020	30/09/2020
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	30/09/2020	30/09/2020
							F12	Serious Injury/ Illness/ Impairment	A070504	Premature Discharge of Battery	30/09/2020	30/09/2020
							F19	Surgical Intervention	A070504	Premature Discharge of Battery	30/09/2020	30/09/2020
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	30/09/2020	30/09/2020
					CT981	Single use	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	30/09/2020	30/09/2020
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	30/09/2020	30/09/2020
							F12	Serious Injury/ Illness/ Impairment	A070504	Premature Discharge of Battery	30/09/2020	30/09/2020
							F19	Surgical Intervention	A070504	Premature Discharge of Battery	30/09/2020	30/09/2020
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	30/09/2020	30/09/2020
					CT983	Surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	30/09/2020	30/09/2020
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	30/09/2020	30/09/2020
							F12	Serious Injury/ Illness/ Impairment	A070504	Premature Discharge of Battery	30/09/2020	30/09/2020
							F19	Surgical Intervention	A070504	Premature Discharge of Battery	30/09/2020	30/09/2020
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	30/09/2020	30/09/2020
10175	2020	All other reportable incidents	37265	Dual-chamber implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A072201	High Impedance		
							F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A072201	High Impedance		
							F26	No Health Consequences or Impact	A072201	High Impedance		
10176	2020	All other reportable incidents	47263	Cardiac resynchronization therapy implantable pacemaker	CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A0908	Incorrect, Inadequate or Imprecise Result or Readings		

							F08	Hospitalization or Prolonged Hospitalization	A0908	Incorrect, Inadequate or Imprecise Result or Readings		
							F1905	Device Revision or Replacement	A0908	Incorrect, Inadequate or Imprecise Result or Readings		
					CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A0908	Incorrect, Inadequate or Imprecise Result or Readings		
							F08	Hospitalization or Prolonged Hospitalization	A0908	Incorrect, Inadequate or Imprecise Result or Readings		
							F1905	Device Revision or Replacement	A0908	Incorrect, Inadequate or Imprecise Result or Readings		
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A0908	Incorrect, Inadequate or Imprecise Result or Readings		
							F08	Hospitalization or Prolonged Hospitalization	A0908	Incorrect, Inadequate or Imprecise Result or Readings		
							F1905	Device Revision or Replacement	A0908	Incorrect, Inadequate or Imprecise Result or Readings		
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A0908	Incorrect, Inadequate or Imprecise Result or Readings		
							F08	Hospitalization or Prolonged Hospitalization	A0908	Incorrect, Inadequate or Imprecise Result or Readings		
							F1905	Device Revision or Replacement	A0908	Incorrect, Inadequate or Imprecise Result or Readings		
10177	2020	All other reportable incidents	35852	Single-chamber implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	16/12/2020	16/12/2020
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	16/12/2020	16/12/2020
							F12	Serious Injury/ Illness/ Impairment	A070504	Premature Discharge of Battery	16/12/2020	16/12/2020
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	16/12/2020	16/12/2020
10178	2020	All other reportable incidents	37265	Dual-chamber implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	15/11/2020	15/11/2020
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	15/11/2020	15/11/2020
							F12	Serious Injury/ Illness/ Impairment	A070504	Premature Discharge of Battery	15/11/2020	15/11/2020
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	15/11/2020	15/11/2020
10179	2020	All other reportable incidents	17882	Physiologic-monitoring defibrillation system	CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070908	Failure to Sense		
							F26	No Health Consequences or Impact	A070908	Failure to Sense		
10180	2020	All other reportable incidents	35852	Single-chamber implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A070909	Oversensing		
							F26	No Health Consequences or Impact	A070909	Oversensing		
10181	2020	All other reportable incidents	36241	Pacemaker/defibrillator lead	CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A072201	High Impedance	07/10/2020	07/10/2020
							F1905	Device Revision or Replacement	A072201	High Impedance	07/10/2020	07/10/2020
							F26	No Health Consequences or Impact	A072201	High Impedance	07/10/2020	07/10/2020
10182	2020	All other reportable incidents	35852	Single-chamber implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A072201	High Impedance		
							F1901	Additional Surgery	A072201	High Impedance		
							F1905	Device Revision or Replacement	A072201	High Impedance		
10183	2020	All other reportable incidents	47265	Dual-chamber implantable pacemaker, rate-responsive	CT141	Electrodes/Leads and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070909	Oversensing		
									A090804	False Positive Result		
							F15	Recognized Device or Procedural Complication	A070909	Oversensing		
									A090804	False Positive Result		
					CT145	Cardiovascular devices	E2403	No Clinical Signs, Symptoms or Conditions	A070909	Oversensing		
									A090804	False Positive Result		
							F15	Recognized Device or Procedural Complication	A070909	Oversensing		
									A090804	False Positive Result		
					CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070909	Oversensing		
									A090804	False Positive Result		
							F15	Recognized Device or Procedural Complication	A070909	Oversensing		
									A090804	False Positive Result		
					CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A070909	Oversensing		
									A090804	False Positive Result		
							F15	Recognized Device or Procedural Complication	A070909	Oversensing		
									A090804	False Positive Result		
					CT233	Surgical	E2403	No Clinical Signs, Symptoms or Conditions	A070909	Oversensing		
									A090804	False Positive Result		
							F15	Recognized Device or Procedural Complication	A070909	Oversensing		
									A090804	False Positive Result		
					CT2406	Implantable	E2403	No Clinical Signs, Symptoms or Conditions	A070909	Oversensing		
									A090804	False Positive Result		
							F15	Recognized Device or Procedural Complication	A070909	Oversensing		
									A090804	False Positive Result		
					CT2407	Active implantable	E2403	No Clinical Signs, Symptoms or Conditions	A070909	Oversensing		
									A090804	False Positive Result		
							F15	Recognized Device or Procedural Complication	A070909	Oversensing		
									A090804	False Positive Result		
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A070909	Oversensing		
									A090804	False Positive Result		
							F15	Recognized Device or Procedural Complication	A070909	Oversensing		
									A090804	False Positive Result		
					CT2727	Active implant devices/ancillaries	E2403	No Clinical Signs, Symptoms or Conditions	A070909	Oversensing		
									A090804	False Positive Result		
							F15	Recognized Device or Procedural Complication	A070909	Oversensing		
									A090804	False Positive Result		
					CT2750	Implantable cardiac electrical stimulators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070909	Oversensing		
									A090804	False Positive Result		
							F15	Recognized Device or Procedural Complication	A070909	Oversensing		
									A090804	False Positive Result		
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A070909	Oversensing		
									A090804	False Positive Result		
							F15	Recognized Device or Procedural Complication	A070909	Oversensing		
									A090804	False Positive Result		
					CT2985	Active	E2403	No Clinical Signs, Symptoms or Conditions	A070909	Oversensing		
									A090804	False Positive Result		
							F15	Recognized Device or Procedural Complication	A070909	Oversensing		
									A090804	False Positive Result		
					CT3001	Implantable metal	E2403	No Clinical Signs, Symptoms or Conditions	A070909	Oversensing		
									A090804	False Positive Result		

							F15	Recognized Device or Procedural Complication	A070909	Oversensing		
									A090804	False Positive Result		
					CT3002	Implantable synthetic polymer	E2403	No Clinical Signs, Symptoms or Conditions	A070909	Oversensing		
									A090804	False Positive Result		
							F15	Recognized Device or Procedural Complication	A070909	Oversensing		
									A090804	False Positive Result		
					CT3006	Implantable inorganic material	E2403	No Clinical Signs, Symptoms or Conditions	A070909	Oversensing		
									A090804	False Positive Result		
							F15	Recognized Device or Procedural Complication	A070909	Oversensing		
									A090804	False Positive Result		
					CT314	Electrically powered	E2403	No Clinical Signs, Symptoms or Conditions	A070909	Oversensing		
									A090804	False Positive Result		
							F15	Recognized Device or Procedural Complication	A070909	Oversensing		
									A090804	False Positive Result		
					CT321	Long-term surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A070909	Oversensing		
									A090804	False Positive Result		
							F15	Recognized Device or Procedural Complication	A070909	Oversensing		
									A090804	False Positive Result		
					CT449	Generators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070909	Oversensing		
									A090804	False Positive Result		
							F15	Recognized Device or Procedural Complication	A070909	Oversensing		
									A090804	False Positive Result		
					CT981	Single use	E2403	No Clinical Signs, Symptoms or Conditions	A070909	Oversensing		
									A090804	False Positive Result		
							F15	Recognized Device or Procedural Complication	A070909	Oversensing		
									A090804	False Positive Result		
					CT983	Surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A070909	Oversensing		
									A090804	False Positive Result		
							F15	Recognized Device or Procedural Complication	A070909	Oversensing		
									A090804	False Positive Result		
10184	2020	All other reportable incidents	35852	Single-chamber implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E2104	Electric Shock	A071302	Inappropriate/Inadequate Shock/Stimulation		
							E2403	No Clinical Signs, Symptoms or Conditions	A071302	Inappropriate/Inadequate Shock/Stimulation		
							F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A071302	Inappropriate/Inadequate Shock/Stimulation		
							F26	No Health Consequences or Impact	A071302	Inappropriate/Inadequate Shock/Stimulation		
10185	2020	All other reportable incidents	47270	Cardiac resynchronization therapy implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A070910	Undersensing		
									A0712	Pacing Problem		
							F26	No Health Consequences or Impact	A070910	Undersensing		
									A0712	Pacing Problem		
10186	2020	All other reportable incidents	35852	Single-chamber implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	06/10/2020	06/10/2020
							F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A070504	Premature Discharge of Battery	06/10/2020	06/10/2020
							F1901	Additional Surgery	A070504	Premature Discharge of Battery	06/10/2020	06/10/2020
10187	2020	All other reportable incidents	47267	Single-chamber implantable pacemaker, rate-responsive	CT141	Electrodes/Leads and associated devices	E0601	Arrhythmia	A070909	Oversensing	07/10/2020	07/10/2020
									A071204	No Pacing	07/10/2020	07/10/2020
									A1205	Loose or Intermittent Connection	07/10/2020	07/10/2020
							F1901	Additional Surgery	A070909	Oversensing	07/10/2020	07/10/2020
									A071204	No Pacing	07/10/2020	07/10/2020
									A1205	Loose or Intermittent Connection	07/10/2020	07/10/2020
							F1905	Device Revision or Replacement	A070909	Oversensing	07/10/2020	07/10/2020
									A071204	No Pacing	07/10/2020	07/10/2020
									A1205	Loose or Intermittent Connection	07/10/2020	07/10/2020
					CT145	Cardiovascular devices	E0601	Arrhythmia	A070909	Oversensing	07/10/2020	07/10/2020
									A071204	No Pacing	07/10/2020	07/10/2020
									A1205	Loose or Intermittent Connection	07/10/2020	07/10/2020
							F1901	Additional Surgery	A070909	Oversensing	07/10/2020	07/10/2020
									A071204	No Pacing	07/10/2020	07/10/2020
									A1205	Loose or Intermittent Connection	07/10/2020	07/10/2020
							F1905	Device Revision or Replacement	A070909	Oversensing	07/10/2020	07/10/2020
									A071204	No Pacing	07/10/2020	07/10/2020
									A1205	Loose or Intermittent Connection	07/10/2020	07/10/2020
					CT1684	Cardiac pacemakers and associated devices	E0601	Arrhythmia	A070909	Oversensing	07/10/2020	07/10/2020
									A071204	No Pacing	07/10/2020	07/10/2020
									A1205	Loose or Intermittent Connection	07/10/2020	07/10/2020
							F1901	Additional Surgery	A070909	Oversensing	07/10/2020	07/10/2020
									A071204	No Pacing	07/10/2020	07/10/2020
									A1205	Loose or Intermittent Connection	07/10/2020	07/10/2020
							F1905	Device Revision or Replacement	A070909	Oversensing	07/10/2020	07/10/2020
									A071204	No Pacing	07/10/2020	07/10/2020
									A1205	Loose or Intermittent Connection	07/10/2020	07/10/2020
					CT1685	Cardiac pacemakers	E0601	Arrhythmia	A070909	Oversensing	07/10/2020	07/10/2020
									A071204	No Pacing	07/10/2020	07/10/2020
									A1205	Loose or Intermittent Connection	07/10/2020	07/10/2020
							F1901	Additional Surgery	A070909	Oversensing	07/10/2020	07/10/2020
									A071204	No Pacing	07/10/2020	07/10/2020
									A1205	Loose or Intermittent Connection	07/10/2020	07/10/2020
							F1905	Device Revision or Replacement	A070909	Oversensing	07/10/2020	07/10/2020
									A071204	No Pacing	07/10/2020	07/10/2020
									A1205	Loose or Intermittent Connection	07/10/2020	07/10/2020
					CT233	Surgical	E0601	Arrhythmia	A070909	Oversensing	07/10/2020	07/10/2020
									A071204	No Pacing	07/10/2020	07/10/2020
									A1205	Loose or Intermittent Connection	07/10/2020	07/10/2020
							F1901	Additional Surgery	A070909	Oversensing	07/10/2020	07/10/2020
									A071204	No Pacing	07/10/2020	07/10/2020
									A1205	Loose or Intermittent Connection	07/10/2020	07/10/2020
							F1905	Device Revision or Replacement	A070909	Oversensing	07/10/2020	07/10/2020
									A071204	No Pacing	07/10/2020	07/10/2020
									A1205	Loose or Intermittent Connection	07/10/2020	07/10/2020
					CT2406	Implantable	E0601	Arrhythmia	A070909	Oversensing	07/10/2020	07/10/2020
									A071204	No Pacing	07/10/2020	07/10/2020

								A1205	Loose or Intermittent Connection	07/10/2020	07/10/2020
						F1901	Additional Surgery	A070909	Oversensing	07/10/2020	07/10/2020
								A071204	No Pacing	07/10/2020	07/10/2020
								A1205	Loose or Intermittent Connection	07/10/2020	07/10/2020
						F1905	Device Revision or Replacement	A070909	Oversensing	07/10/2020	07/10/2020
								A071204	No Pacing	07/10/2020	07/10/2020
								A1205	Loose or Intermittent Connection	07/10/2020	07/10/2020
					CT2407	Active implantable	E0601	Arrhythmia	A070909	Oversensing	07/10/2020	07/10/2020
									A071204	No Pacing	07/10/2020	07/10/2020
									A1205	Loose or Intermittent Connection	07/10/2020	07/10/2020
							F1901	Additional Surgery	A070909	Oversensing	07/10/2020	07/10/2020
									A071204	No Pacing	07/10/2020	07/10/2020
									A1205	Loose or Intermittent Connection	07/10/2020	07/10/2020
							F1905	Device Revision or Replacement	A070909	Oversensing	07/10/2020	07/10/2020
									A071204	No Pacing	07/10/2020	07/10/2020
									A1205	Loose or Intermittent Connection	07/10/2020	07/10/2020
					CT257	Implantable cardiac pacemakers	E0601	Arrhythmia	A070909	Oversensing	07/10/2020	07/10/2020
									A071204	No Pacing	07/10/2020	07/10/2020
									A1205	Loose or Intermittent Connection	07/10/2020	07/10/2020
							F1901	Additional Surgery	A070909	Oversensing	07/10/2020	07/10/2020
									A071204	No Pacing	07/10/2020	07/10/2020
									A1205	Loose or Intermittent Connection	07/10/2020	07/10/2020
							F1905	Device Revision or Replacement	A070909	Oversensing	07/10/2020	07/10/2020
									A071204	No Pacing	07/10/2020	07/10/2020
									A1205	Loose or Intermittent Connection	07/10/2020	07/10/2020
					CT2727	Active implant devices/ancillaries	E0601	Arrhythmia	A070909	Oversensing	07/10/2020	07/10/2020
									A071204	No Pacing	07/10/2020	07/10/2020
									A1205	Loose or Intermittent Connection	07/10/2020	07/10/2020
							F1901	Additional Surgery	A070909	Oversensing	07/10/2020	07/10/2020
									A071204	No Pacing	07/10/2020	07/10/2020
									A1205	Loose or Intermittent Connection	07/10/2020	07/10/2020
							F1905	Device Revision or Replacement	A070909	Oversensing	07/10/2020	07/10/2020
									A071204	No Pacing	07/10/2020	07/10/2020
									A1205	Loose or Intermittent Connection	07/10/2020	07/10/2020
					CT2750	Implantable cardiac electrical stimulators and associated devices	E0601	Arrhythmia	A070909	Oversensing	07/10/2020	07/10/2020
									A071204	No Pacing	07/10/2020	07/10/2020
									A1205	Loose or Intermittent Connection	07/10/2020	07/10/2020
							F1901	Additional Surgery	A070909	Oversensing	07/10/2020	07/10/2020
									A071204	No Pacing	07/10/2020	07/10/2020
									A1205	Loose or Intermittent Connection	07/10/2020	07/10/2020
							F1905	Device Revision or Replacement	A070909	Oversensing	07/10/2020	07/10/2020
									A071204	No Pacing	07/10/2020	07/10/2020
									A1205	Loose or Intermittent Connection	07/10/2020	07/10/2020
					CT2866	Implantable cardiac electrical stimulators	E0601	Arrhythmia	A070909	Oversensing	07/10/2020	07/10/2020
									A071204	No Pacing	07/10/2020	07/10/2020
									A1205	Loose or Intermittent Connection	07/10/2020	07/10/2020
							F1901	Additional Surgery	A070909	Oversensing	07/10/2020	07/10/2020
									A071204	No Pacing	07/10/2020	07/10/2020
									A1205	Loose or Intermittent Connection	07/10/2020	07/10/2020
							F1905	Device Revision or Replacement	A070909	Oversensing	07/10/2020	07/10/2020
									A071204	No Pacing	07/10/2020	07/10/2020
									A1205	Loose or Intermittent Connection	07/10/2020	07/10/2020
					CT2985	Active	E0601	Arrhythmia	A070909	Oversensing	07/10/2020	07/10/2020
									A071204	No Pacing	07/10/2020	07/10/2020
									A1205	Loose or Intermittent Connection	07/10/2020	07/10/2020
							F1901	Additional Surgery	A070909	Oversensing	07/10/2020	07/10/2020
									A071204	No Pacing	07/10/2020	07/10/2020
									A1205	Loose or Intermittent Connection	07/10/2020	07/10/2020
							F1905	Device Revision or Replacement	A070909	Oversensing	07/10/2020	07/10/2020
									A071204	No Pacing	07/10/2020	07/10/2020
									A1205	Loose or Intermittent Connection	07/10/2020	07/10/2020
					CT3001	Implantable metal	E0601	Arrhythmia	A070909	Oversensing	07/10/2020	07/10/2020
									A071204	No Pacing	07/10/2020	07/10/2020
									A1205	Loose or Intermittent Connection	07/10/2020	07/10/2020
							F1901	Additional Surgery	A070909	Oversensing	07/10/2020	07/10/2020
									A071204	No Pacing	07/10/2020	07/10/2020
									A1205	Loose or Intermittent Connection	07/10/2020	07/10/2020
							F1905	Device Revision or Replacement	A070909	Oversensing	07/10/2020	07/10/2020
									A071204	No Pacing	07/10/2020	07/10/2020
									A1205	Loose or Intermittent Connection	07/10/2020	07/10/2020
					CT3002	Implantable synthetic polymer	E0601	Arrhythmia	A070909	Oversensing	07/10/2020	07/10/2020
									A071204	No Pacing	07/10/2020	07/10/2020
									A1205	Loose or Intermittent Connection	07/10/2020	07/10/2020
							F1901	Additional Surgery	A070909	Oversensing	07/10/2020	07/10/2020
									A071204	No Pacing	07/10/2020	07/10/2020
									A1205	Loose or Intermittent Connection	07/10/2020	07/10/2020
							F1905	Device Revision or Replacement	A070909	Oversensing	07/10/2020	07/10/2020
									A071204	No Pacing	07/10/2020	07/10/2020
									A1205	Loose or Intermittent Connection	07/10/2020	07/10/2020
					CT3006	Implantable inorganic material	E0601	Arrhythmia	A070909	Oversensing	07/10/2020	07/10/2020
									A071204	No Pacing	07/10/2020	07/10/2020
									A1205	Loose or Intermittent Connection	07/10/2020	07/10/2020
							F1901	Additional Surgery	A070909	Oversensing	07/10/2020	07/10/2020
									A071204	No Pacing	07/10/2020	07/10/2020
									A1205	Loose or Intermittent Connection	07/10/2020	07/10/2020
							F1905	Device Revision or Replacement	A070909	Oversensing	07/10/2020	07/10/2020
									A071204	No Pacing	07/10/2020	07/10/2020
									A1205	Loose or Intermittent Connection	07/10/2020	07/10/2020

					CT314	Electrically powered	E0601	Arrhythmia	A070909	Oversensing	07/10/2020	07/10/2020
									A071204	No Pacing	07/10/2020	07/10/2020
									A1205	Loose or Intermittent Connection	07/10/2020	07/10/2020
							F1901	Additional Surgery	A070909	Oversensing	07/10/2020	07/10/2020
									A071204	No Pacing	07/10/2020	07/10/2020
									A1205	Loose or Intermittent Connection	07/10/2020	07/10/2020
							F1905	Device Revision or Replacement	A070909	Oversensing	07/10/2020	07/10/2020
									A071204	No Pacing	07/10/2020	07/10/2020
									A1205	Loose or Intermittent Connection	07/10/2020	07/10/2020
					CT321	Long-term surgical invasive	E0601	Arrhythmia	A070909	Oversensing	07/10/2020	07/10/2020
									A071204	No Pacing	07/10/2020	07/10/2020
									A1205	Loose or Intermittent Connection	07/10/2020	07/10/2020
							F1901	Additional Surgery	A070909	Oversensing	07/10/2020	07/10/2020
									A071204	No Pacing	07/10/2020	07/10/2020
									A1205	Loose or Intermittent Connection	07/10/2020	07/10/2020
							F1905	Device Revision or Replacement	A070909	Oversensing	07/10/2020	07/10/2020
									A071204	No Pacing	07/10/2020	07/10/2020
									A1205	Loose or Intermittent Connection	07/10/2020	07/10/2020
					CT449	Generators and associated devices	E0601	Arrhythmia	A070909	Oversensing	07/10/2020	07/10/2020
									A071204	No Pacing	07/10/2020	07/10/2020
									A1205	Loose or Intermittent Connection	07/10/2020	07/10/2020
							F1901	Additional Surgery	A070909	Oversensing	07/10/2020	07/10/2020
									A071204	No Pacing	07/10/2020	07/10/2020
									A1205	Loose or Intermittent Connection	07/10/2020	07/10/2020
							F1905	Device Revision or Replacement	A070909	Oversensing	07/10/2020	07/10/2020
									A071204	No Pacing	07/10/2020	07/10/2020
									A1205	Loose or Intermittent Connection	07/10/2020	07/10/2020
					CT981	Single use	E0601	Arrhythmia	A070909	Oversensing	07/10/2020	07/10/2020
									A071204	No Pacing	07/10/2020	07/10/2020
									A1205	Loose or Intermittent Connection	07/10/2020	07/10/2020
							F1901	Additional Surgery	A070909	Oversensing	07/10/2020	07/10/2020
									A071204	No Pacing	07/10/2020	07/10/2020
									A1205	Loose or Intermittent Connection	07/10/2020	07/10/2020
							F1905	Device Revision or Replacement	A070909	Oversensing	07/10/2020	07/10/2020
									A071204	No Pacing	07/10/2020	07/10/2020
									A1205	Loose or Intermittent Connection	07/10/2020	07/10/2020
					CT983	Surgical invasive	E0601	Arrhythmia	A070909	Oversensing	07/10/2020	07/10/2020
									A071204	No Pacing	07/10/2020	07/10/2020
									A1205	Loose or Intermittent Connection	07/10/2020	07/10/2020
							F1901	Additional Surgery	A070909	Oversensing	07/10/2020	07/10/2020
									A071204	No Pacing	07/10/2020	07/10/2020
									A1205	Loose or Intermittent Connection	07/10/2020	07/10/2020
							F1905	Device Revision or Replacement	A070909	Oversensing	07/10/2020	07/10/2020
									A071204	No Pacing	07/10/2020	07/10/2020
									A1205	Loose or Intermittent Connection	07/10/2020	07/10/2020
10188	2020	All other reportable incidents	35223	Endocardial/interventricular septal pacing lead	CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A0401	Break	29/04/2021	29/04/2021
									A070909	Oversensing	29/04/2021	29/04/2021
									A090801	Signal Artifact/Noise	29/04/2021	29/04/2021
							F26	No Health Consequences or Impact	A0401	Break	29/04/2021	29/04/2021
									A070909	Oversensing	29/04/2021	29/04/2021
									A090801	Signal Artifact/Noise	29/04/2021	29/04/2021
10189	2020	All other reportable incidents	47265	Dual-chamber implantable pacemaker, rate-responsive	CT141	Electrodes/Leads and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	13/01/2021	13/01/2021
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	13/01/2021	13/01/2021
							F12	Serious Injury/ Illness/ Impairment	A070504	Premature Discharge of Battery	13/01/2021	13/01/2021
							F1903	Device Explantation	A070504	Premature Discharge of Battery	13/01/2021	13/01/2021
					CT145	Cardiovascular devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	13/01/2021	13/01/2021
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	13/01/2021	13/01/2021
							F12	Serious Injury/ Illness/ Impairment	A070504	Premature Discharge of Battery	13/01/2021	13/01/2021
							F1903	Device Explantation	A070504	Premature Discharge of Battery	13/01/2021	13/01/2021
					CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	13/01/2021	13/01/2021
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	13/01/2021	13/01/2021
							F12	Serious Injury/ Illness/ Impairment	A070504	Premature Discharge of Battery	13/01/2021	13/01/2021
							F1903	Device Explantation	A070504	Premature Discharge of Battery	13/01/2021	13/01/2021
					CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	13/01/2021	13/01/2021
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	13/01/2021	13/01/2021
							F12	Serious Injury/ Illness/ Impairment	A070504	Premature Discharge of Battery	13/01/2021	13/01/2021
							F1903	Device Explantation	A070504	Premature Discharge of Battery	13/01/2021	13/01/2021
					CT233	Surgical	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	13/01/2021	13/01/2021
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	13/01/2021	13/01/2021
							F12	Serious Injury/ Illness/ Impairment	A070504	Premature Discharge of Battery	13/01/2021	13/01/2021
							F1903	Device Explantation	A070504	Premature Discharge of Battery	13/01/2021	13/01/2021
					CT2406	Implantable	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	13/01/2021	13/01/2021
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	13/01/2021	13/01/2021
							F12	Serious Injury/ Illness/ Impairment	A070504	Premature Discharge of Battery	13/01/2021	13/01/2021
							F1903	Device Explantation	A070504	Premature Discharge of Battery	13/01/2021	13/01/2021
					CT2407	Active implantable	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	13/01/2021	13/01/2021
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	13/01/2021	13/01/2021
							F12	Serious Injury/ Illness/ Impairment	A070504	Premature Discharge of Battery	13/01/2021	13/01/2021
							F1903	Device Explantation	A070504	Premature Discharge of Battery	13/01/2021	13/01/2021
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	13/01/2021	13/01/2021
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	13/01/2021	13/01/2021
							F12	Serious Injury/ Illness/ Impairment	A070504	Premature Discharge of Battery	13/01/2021	13/01/2021
							F1903	Device Explantation	A070504	Premature Discharge of Battery	13/01/2021	13/01/2021
					CT2727	Active implant devices/ancillaries	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	13/01/2021	13/01/2021
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	13/01/2021	13/01/2021
							F12	Serious Injury/ Illness/ Impairment	A070504	Premature Discharge of Battery	13/01/2021	13/01/2021
							F1903	Device Explantation	A070504	Premature Discharge of Battery	13/01/2021	13/01/2021
					CT2750	Implantable cardiac electrical stimulators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	13/01/2021	13/01/2021

							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	13/01/2021	13/01/2021
							F12	Serious Injury/ Illness/ Impairment	A070504	Premature Discharge of Battery	13/01/2021	13/01/2021
							F1903	Device Explantation	A070504	Premature Discharge of Battery	13/01/2021	13/01/2021
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	13/01/2021	13/01/2021
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	13/01/2021	13/01/2021
							F12	Serious Injury/ Illness/ Impairment	A070504	Premature Discharge of Battery	13/01/2021	13/01/2021
							F1903	Device Explantation	A070504	Premature Discharge of Battery	13/01/2021	13/01/2021
					CT2985	Active	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	13/01/2021	13/01/2021
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	13/01/2021	13/01/2021
							F12	Serious Injury/ Illness/ Impairment	A070504	Premature Discharge of Battery	13/01/2021	13/01/2021
							F1903	Device Explantation	A070504	Premature Discharge of Battery	13/01/2021	13/01/2021
					CT3001	Implantable metal	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	13/01/2021	13/01/2021
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	13/01/2021	13/01/2021
							F12	Serious Injury/ Illness/ Impairment	A070504	Premature Discharge of Battery	13/01/2021	13/01/2021
							F1903	Device Explantation	A070504	Premature Discharge of Battery	13/01/2021	13/01/2021
					CT3002	Implantable synthetic polymer	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	13/01/2021	13/01/2021
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	13/01/2021	13/01/2021
							F12	Serious Injury/ Illness/ Impairment	A070504	Premature Discharge of Battery	13/01/2021	13/01/2021
							F1903	Device Explantation	A070504	Premature Discharge of Battery	13/01/2021	13/01/2021
					CT3006	Implantable inorganic material	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	13/01/2021	13/01/2021
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	13/01/2021	13/01/2021
							F12	Serious Injury/ Illness/ Impairment	A070504	Premature Discharge of Battery	13/01/2021	13/01/2021
							F1903	Device Explantation	A070504	Premature Discharge of Battery	13/01/2021	13/01/2021
					CT314	Electrically powered	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	13/01/2021	13/01/2021
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	13/01/2021	13/01/2021
							F12	Serious Injury/ Illness/ Impairment	A070504	Premature Discharge of Battery	13/01/2021	13/01/2021
							F1903	Device Explantation	A070504	Premature Discharge of Battery	13/01/2021	13/01/2021
					CT321	Long-term surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	13/01/2021	13/01/2021
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	13/01/2021	13/01/2021
							F12	Serious Injury/ Illness/ Impairment	A070504	Premature Discharge of Battery	13/01/2021	13/01/2021
							F1903	Device Explantation	A070504	Premature Discharge of Battery	13/01/2021	13/01/2021
					CT449	Generators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	13/01/2021	13/01/2021
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	13/01/2021	13/01/2021
							F12	Serious Injury/ Illness/ Impairment	A070504	Premature Discharge of Battery	13/01/2021	13/01/2021
							F1903	Device Explantation	A070504	Premature Discharge of Battery	13/01/2021	13/01/2021
					CT981	Single use	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	13/01/2021	13/01/2021
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	13/01/2021	13/01/2021
							F12	Serious Injury/ Illness/ Impairment	A070504	Premature Discharge of Battery	13/01/2021	13/01/2021
							F1903	Device Explantation	A070504	Premature Discharge of Battery	13/01/2021	13/01/2021
					CT983	Surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	13/01/2021	13/01/2021
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	13/01/2021	13/01/2021
							F12	Serious Injury/ Illness/ Impairment	A070504	Premature Discharge of Battery	13/01/2021	13/01/2021
							F1903	Device Explantation	A070504	Premature Discharge of Battery	13/01/2021	13/01/2021
10190	2020	All other reportable incidents	47267	Single-chamber implantable pacemaker, rate-responsive	CT141	Electrodes/Leads and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	19/10/2020	19/10/2020
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	19/10/2020	19/10/2020
							F12	Serious Injury/ Illness/ Impairment	A070504	Premature Discharge of Battery	19/10/2020	19/10/2020
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	19/10/2020	19/10/2020
					CT145	Cardiovascular devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	19/10/2020	19/10/2020
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	19/10/2020	19/10/2020
							F12	Serious Injury/ Illness/ Impairment	A070504	Premature Discharge of Battery	19/10/2020	19/10/2020
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	19/10/2020	19/10/2020
					CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	19/10/2020	19/10/2020
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	19/10/2020	19/10/2020
							F12	Serious Injury/ Illness/ Impairment	A070504	Premature Discharge of Battery	19/10/2020	19/10/2020
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	19/10/2020	19/10/2020
					CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	19/10/2020	19/10/2020
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	19/10/2020	19/10/2020
							F12	Serious Injury/ Illness/ Impairment	A070504	Premature Discharge of Battery	19/10/2020	19/10/2020
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	19/10/2020	19/10/2020
					CT233	Surgical	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	19/10/2020	19/10/2020
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	19/10/2020	19/10/2020
							F12	Serious Injury/ Illness/ Impairment	A070504	Premature Discharge of Battery	19/10/2020	19/10/2020
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	19/10/2020	19/10/2020
					CT2406	Implantable	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	19/10/2020	19/10/2020
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	19/10/2020	19/10/2020
							F12	Serious Injury/ Illness/ Impairment	A070504	Premature Discharge of Battery	19/10/2020	19/10/2020
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	19/10/2020	19/10/2020
					CT2407	Active implantable	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	19/10/2020	19/10/2020
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	19/10/2020	19/10/2020
							F12	Serious Injury/ Illness/ Impairment	A070504	Premature Discharge of Battery	19/10/2020	19/10/2020
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	19/10/2020	19/10/2020
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	19/10/2020	19/10/2020
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	19/10/2020	19/10/2020
							F12	Serious Injury/ Illness/ Impairment	A070504	Premature Discharge of Battery	19/10/2020	19/10/2020
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	19/10/2020	19/10/2020
					CT2727	Active implant devices/ancillaries	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	19/10/2020	19/10/2020
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	19/10/2020	19/10/2020
							F12	Serious Injury/ Illness/ Impairment	A070504	Premature Discharge of Battery	19/10/2020	19/10/2020
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	19/10/2020	19/10/2020
					CT2750	Implantable cardiac electrical stimulators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	19/10/2020	19/10/2020
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	19/10/2020	19/10/2020
							F12	Serious Injury/ Illness/ Impairment	A070504	Premature Discharge of Battery	19/10/2020	19/10/2020
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	19/10/2020	19/10/2020
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	19/10/2020	19/10/2020
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	19/10/2020	19/10/2020
							F12	Serious Injury/ Illness/ Impairment	A070504	Premature Discharge of Battery	19/10/2020	19/10/2020
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	19/10/2020	19/10/2020
					CT2985	Active	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	19/10/2020	19/10/2020

							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	19/10/2020	19/10/2020
							F12	Serious Injury/ Illness/ Impairment	A070504	Premature Discharge of Battery	19/10/2020	19/10/2020
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	19/10/2020	19/10/2020
					CT3001	Implantable metal	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	19/10/2020	19/10/2020
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	19/10/2020	19/10/2020
							F12	Serious Injury/ Illness/ Impairment	A070504	Premature Discharge of Battery	19/10/2020	19/10/2020
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	19/10/2020	19/10/2020
					CT3002	Implantable synthetic polymer	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	19/10/2020	19/10/2020
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	19/10/2020	19/10/2020
							F12	Serious Injury/ Illness/ Impairment	A070504	Premature Discharge of Battery	19/10/2020	19/10/2020
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	19/10/2020	19/10/2020
					CT3006	Implantable inorganic material	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	19/10/2020	19/10/2020
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	19/10/2020	19/10/2020
							F12	Serious Injury/ Illness/ Impairment	A070504	Premature Discharge of Battery	19/10/2020	19/10/2020
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	19/10/2020	19/10/2020
					CT314	Electrically powered	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	19/10/2020	19/10/2020
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	19/10/2020	19/10/2020
							F12	Serious Injury/ Illness/ Impairment	A070504	Premature Discharge of Battery	19/10/2020	19/10/2020
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	19/10/2020	19/10/2020
					CT321	Long-term surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	19/10/2020	19/10/2020
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	19/10/2020	19/10/2020
							F12	Serious Injury/ Illness/ Impairment	A070504	Premature Discharge of Battery	19/10/2020	19/10/2020
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	19/10/2020	19/10/2020
					CT449	Generators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	19/10/2020	19/10/2020
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	19/10/2020	19/10/2020
							F12	Serious Injury/ Illness/ Impairment	A070504	Premature Discharge of Battery	19/10/2020	19/10/2020
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	19/10/2020	19/10/2020
					CT981	Single use	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	19/10/2020	19/10/2020
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	19/10/2020	19/10/2020
							F12	Serious Injury/ Illness/ Impairment	A070504	Premature Discharge of Battery	19/10/2020	19/10/2020
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	19/10/2020	19/10/2020
					CT983	Surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	19/10/2020	19/10/2020
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	19/10/2020	19/10/2020
							F12	Serious Injury/ Illness/ Impairment	A070504	Premature Discharge of Battery	19/10/2020	19/10/2020
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	19/10/2020	19/10/2020
10191	2020	All other reportable incidents	47263	Cardiac resynchronization therapy implantable pacemaker	CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	23/11/2020	23/11/2020
							F19	Surgical Intervention	A070504	Premature Discharge of Battery	23/11/2020	23/11/2020
					CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	23/11/2020	23/11/2020
							F19	Surgical Intervention	A070504	Premature Discharge of Battery	23/11/2020	23/11/2020
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	23/11/2020	23/11/2020
							F19	Surgical Intervention	A070504	Premature Discharge of Battery	23/11/2020	23/11/2020
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	23/11/2020	23/11/2020
							F19	Surgical Intervention	A070504	Premature Discharge of Battery	23/11/2020	23/11/2020
10192	2020	All other reportable incidents	47265	Dual-chamber implantable pacemaker, rate-responsive	CT141	Electrodes/Leads and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	19/12/2020	19/12/2020
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	19/12/2020	19/12/2020
							F12	Serious Injury/ Illness/ Impairment	A070504	Premature Discharge of Battery	19/12/2020	19/12/2020
							F19	Surgical Intervention	A070504	Premature Discharge of Battery	19/12/2020	19/12/2020
					CT145	Cardiovascular devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	19/12/2020	19/12/2020
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	19/12/2020	19/12/2020
							F12	Serious Injury/ Illness/ Impairment	A070504	Premature Discharge of Battery	19/12/2020	19/12/2020
							F19	Surgical Intervention	A070504	Premature Discharge of Battery	19/12/2020	19/12/2020
					CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	19/12/2020	19/12/2020
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	19/12/2020	19/12/2020
							F12	Serious Injury/ Illness/ Impairment	A070504	Premature Discharge of Battery	19/12/2020	19/12/2020
							F19	Surgical Intervention	A070504	Premature Discharge of Battery	19/12/2020	19/12/2020
					CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	19/12/2020	19/12/2020
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	19/12/2020	19/12/2020
							F12	Serious Injury/ Illness/ Impairment	A070504	Premature Discharge of Battery	19/12/2020	19/12/2020
							F19	Surgical Intervention	A070504	Premature Discharge of Battery	19/12/2020	19/12/2020
					CT233	Surgical	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	19/12/2020	19/12/2020
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	19/12/2020	19/12/2020
							F12	Serious Injury/ Illness/ Impairment	A070504	Premature Discharge of Battery	19/12/2020	19/12/2020
							F19	Surgical Intervention	A070504	Premature Discharge of Battery	19/12/2020	19/12/2020
					CT2406	Implantable	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	19/12/2020	19/12/2020
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	19/12/2020	19/12/2020
							F12	Serious Injury/ Illness/ Impairment	A070504	Premature Discharge of Battery	19/12/2020	19/12/2020
							F19	Surgical Intervention	A070504	Premature Discharge of Battery	19/12/2020	19/12/2020
					CT2407	Active implantable	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	19/12/2020	19/12/2020
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	19/12/2020	19/12/2020
							F12	Serious Injury/ Illness/ Impairment	A070504	Premature Discharge of Battery	19/12/2020	19/12/2020
							F19	Surgical Intervention	A070504	Premature Discharge of Battery	19/12/2020	19/12/2020
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	19/12/2020	19/12/2020
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	19/12/2020	19/12/2020
							F12	Serious Injury/ Illness/ Impairment	A070504	Premature Discharge of Battery	19/12/2020	19/12/2020
							F19	Surgical Intervention	A070504	Premature Discharge of Battery	19/12/2020	19/12/2020
					CT2727	Active implant devices/ancillaries	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	19/12/2020	19/12/2020
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	19/12/2020	19/12/2020
							F12	Serious Injury/ Illness/ Impairment	A070504	Premature Discharge of Battery	19/12/2020	19/12/2020
							F19	Surgical Intervention	A070504	Premature Discharge of Battery	19/12/2020	19/12/2020
					CT2750	Implantable cardiac electrical stimulators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	19/12/2020	19/12/2020
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	19/12/2020	19/12/2020
							F12	Serious Injury/ Illness/ Impairment	A070504	Premature Discharge of Battery	19/12/2020	19/12/2020
							F19	Surgical Intervention	A070504	Premature Discharge of Battery	19/12/2020	19/12/2020
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	19/12/2020	19/12/2020
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	19/12/2020	19/12/2020
							F12	Serious Injury/ Illness/ Impairment	A070504	Premature Discharge of Battery	19/12/2020	19/12/2020
							F19	Surgical Intervention	A070504	Premature Discharge of Battery	19/12/2020	19/12/2020
					CT2985	Active	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	19/12/2020	19/12/2020

							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	19/12/2020	19/12/2020
							F12	Serious Injury/ Illness/ Impairment	A070504	Premature Discharge of Battery	19/12/2020	19/12/2020
							F19	Surgical Intervention	A070504	Premature Discharge of Battery	19/12/2020	19/12/2020
					CT3001	Implantable metal	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	19/12/2020	19/12/2020
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	19/12/2020	19/12/2020
							F12	Serious Injury/ Illness/ Impairment	A070504	Premature Discharge of Battery	19/12/2020	19/12/2020
							F19	Surgical Intervention	A070504	Premature Discharge of Battery	19/12/2020	19/12/2020
					CT3002	Implantable synthetic polymer	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	19/12/2020	19/12/2020
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	19/12/2020	19/12/2020
							F12	Serious Injury/ Illness/ Impairment	A070504	Premature Discharge of Battery	19/12/2020	19/12/2020
							F19	Surgical Intervention	A070504	Premature Discharge of Battery	19/12/2020	19/12/2020
					CT3006	Implantable inorganic material	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	19/12/2020	19/12/2020
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	19/12/2020	19/12/2020
							F12	Serious Injury/ Illness/ Impairment	A070504	Premature Discharge of Battery	19/12/2020	19/12/2020
							F19	Surgical Intervention	A070504	Premature Discharge of Battery	19/12/2020	19/12/2020
					CT314	Electrically powered	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	19/12/2020	19/12/2020
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	19/12/2020	19/12/2020
							F12	Serious Injury/ Illness/ Impairment	A070504	Premature Discharge of Battery	19/12/2020	19/12/2020
							F19	Surgical Intervention	A070504	Premature Discharge of Battery	19/12/2020	19/12/2020
					CT321	Long-term surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	19/12/2020	19/12/2020
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	19/12/2020	19/12/2020
							F12	Serious Injury/ Illness/ Impairment	A070504	Premature Discharge of Battery	19/12/2020	19/12/2020
							F19	Surgical Intervention	A070504	Premature Discharge of Battery	19/12/2020	19/12/2020
					CT449	Generators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	19/12/2020	19/12/2020
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	19/12/2020	19/12/2020
							F12	Serious Injury/ Illness/ Impairment	A070504	Premature Discharge of Battery	19/12/2020	19/12/2020
							F19	Surgical Intervention	A070504	Premature Discharge of Battery	19/12/2020	19/12/2020
					CT981	Single use	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	19/12/2020	19/12/2020
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	19/12/2020	19/12/2020
							F12	Serious Injury/ Illness/ Impairment	A070504	Premature Discharge of Battery	19/12/2020	19/12/2020
							F19	Surgical Intervention	A070504	Premature Discharge of Battery	19/12/2020	19/12/2020
					CT983	Surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	19/12/2020	19/12/2020
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	19/12/2020	19/12/2020
							F12	Serious Injury/ Illness/ Impairment	A070504	Premature Discharge of Battery	19/12/2020	19/12/2020
							F19	Surgical Intervention	A070504	Premature Discharge of Battery	19/12/2020	19/12/2020
10193	2020	All other reportable incidents	47265	Dual-chamber implantable pacemaker, rate-responsive	CT141	Electrodes/Leads and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	13/01/2021	13/01/2021
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	13/01/2021	13/01/2021
							F12	Serious Injury/ Illness/ Impairment	A070504	Premature Discharge of Battery	13/01/2021	13/01/2021
							F19	Surgical Intervention	A070504	Premature Discharge of Battery	13/01/2021	13/01/2021
					CT145	Cardiovascular devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	13/01/2021	13/01/2021
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	13/01/2021	13/01/2021
							F12	Serious Injury/ Illness/ Impairment	A070504	Premature Discharge of Battery	13/01/2021	13/01/2021
							F19	Surgical Intervention	A070504	Premature Discharge of Battery	13/01/2021	13/01/2021
					CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	13/01/2021	13/01/2021
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	13/01/2021	13/01/2021
							F12	Serious Injury/ Illness/ Impairment	A070504	Premature Discharge of Battery	13/01/2021	13/01/2021
							F19	Surgical Intervention	A070504	Premature Discharge of Battery	13/01/2021	13/01/2021
					CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	13/01/2021	13/01/2021
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	13/01/2021	13/01/2021
							F12	Serious Injury/ Illness/ Impairment	A070504	Premature Discharge of Battery	13/01/2021	13/01/2021
							F19	Surgical Intervention	A070504	Premature Discharge of Battery	13/01/2021	13/01/2021
					CT233	Surgical	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	13/01/2021	13/01/2021
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	13/01/2021	13/01/2021
							F12	Serious Injury/ Illness/ Impairment	A070504	Premature Discharge of Battery	13/01/2021	13/01/2021
							F19	Surgical Intervention	A070504	Premature Discharge of Battery	13/01/2021	13/01/2021
					CT2406	Implantable	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	13/01/2021	13/01/2021
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	13/01/2021	13/01/2021
							F12	Serious Injury/ Illness/ Impairment	A070504	Premature Discharge of Battery	13/01/2021	13/01/2021
							F19	Surgical Intervention	A070504	Premature Discharge of Battery	13/01/2021	13/01/2021
					CT2407	Active implantable	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	13/01/2021	13/01/2021
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	13/01/2021	13/01/2021
							F12	Serious Injury/ Illness/ Impairment	A070504	Premature Discharge of Battery	13/01/2021	13/01/2021
							F19	Surgical Intervention	A070504	Premature Discharge of Battery	13/01/2021	13/01/2021
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	13/01/2021	13/01/2021
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	13/01/2021	13/01/2021
							F12	Serious Injury/ Illness/ Impairment	A070504	Premature Discharge of Battery	13/01/2021	13/01/2021
							F19	Surgical Intervention	A070504	Premature Discharge of Battery	13/01/2021	13/01/2021
					CT2727	Active implant devices/ancillaries	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	13/01/2021	13/01/2021
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	13/01/2021	13/01/2021
							F12	Serious Injury/ Illness/ Impairment	A070504	Premature Discharge of Battery	13/01/2021	13/01/2021
							F19	Surgical Intervention	A070504	Premature Discharge of Battery	13/01/2021	13/01/2021
					CT2750	Implantable cardiac electrical stimulators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	13/01/2021	13/01/2021
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	13/01/2021	13/01/2021
							F12	Serious Injury/ Illness/ Impairment	A070504	Premature Discharge of Battery	13/01/2021	13/01/2021
							F19	Surgical Intervention	A070504	Premature Discharge of Battery	13/01/2021	13/01/2021
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	13/01/2021	13/01/2021
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	13/01/2021	13/01/2021
							F12	Serious Injury/ Illness/ Impairment	A070504	Premature Discharge of Battery	13/01/2021	13/01/2021
							F19	Surgical Intervention	A070504	Premature Discharge of Battery	13/01/2021	13/01/2021
					CT2985	Active	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	13/01/2021	13/01/2021
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	13/01/2021	13/01/2021
							F12	Serious Injury/ Illness/ Impairment	A070504	Premature Discharge of Battery	13/01/2021	13/01/2021
							F19	Surgical Intervention	A070504	Premature Discharge of Battery	13/01/2021	13/01/2021
					CT3001	Implantable metal	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	13/01/2021	13/01/2021
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	13/01/2021	13/01/2021
							F12	Serious Injury/ Illness/ Impairment	A070504	Premature Discharge of Battery	13/01/2021	13/01/2021
							F19	Surgical Intervention	A070504	Premature Discharge of Battery	13/01/2021	13/01/2021
					CT3002	Implantable synthetic polymer	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	13/01/2021	13/01/2021

							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	13/01/2021	13/01/2021
							F12	Serious Injury/ Illness/ Impairment	A070504	Premature Discharge of Battery	13/01/2021	13/01/2021
							F19	Surgical Intervention	A070504	Premature Discharge of Battery	13/01/2021	13/01/2021
					CT3006	Implantable inorganic material	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	13/01/2021	13/01/2021
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	13/01/2021	13/01/2021
							F12	Serious Injury/ Illness/ Impairment	A070504	Premature Discharge of Battery	13/01/2021	13/01/2021
							F19	Surgical Intervention	A070504	Premature Discharge of Battery	13/01/2021	13/01/2021
					CT314	Electrically powered	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	13/01/2021	13/01/2021
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	13/01/2021	13/01/2021
							F12	Serious Injury/ Illness/ Impairment	A070504	Premature Discharge of Battery	13/01/2021	13/01/2021
							F19	Surgical Intervention	A070504	Premature Discharge of Battery	13/01/2021	13/01/2021
					CT321	Long-term surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	13/01/2021	13/01/2021
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	13/01/2021	13/01/2021
							F12	Serious Injury/ Illness/ Impairment	A070504	Premature Discharge of Battery	13/01/2021	13/01/2021
							F19	Surgical Intervention	A070504	Premature Discharge of Battery	13/01/2021	13/01/2021
					CT449	Generators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	13/01/2021	13/01/2021
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	13/01/2021	13/01/2021
							F12	Serious Injury/ Illness/ Impairment	A070504	Premature Discharge of Battery	13/01/2021	13/01/2021
							F19	Surgical Intervention	A070504	Premature Discharge of Battery	13/01/2021	13/01/2021
					CT981	Single use	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	13/01/2021	13/01/2021
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	13/01/2021	13/01/2021
							F12	Serious Injury/ Illness/ Impairment	A070504	Premature Discharge of Battery	13/01/2021	13/01/2021
							F19	Surgical Intervention	A070504	Premature Discharge of Battery	13/01/2021	13/01/2021
					CT983	Surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	13/01/2021	13/01/2021
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	13/01/2021	13/01/2021
							F12	Serious Injury/ Illness/ Impairment	A070504	Premature Discharge of Battery	13/01/2021	13/01/2021
							F19	Surgical Intervention	A070504	Premature Discharge of Battery	13/01/2021	13/01/2021
10194	2020	All other reportable incidents	35852	Single-chamber implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	07/10/2020	07/10/2020
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	07/10/2020	07/10/2020
10195	2020	All other reportable incidents	47265	Dual-chamber implantable pacemaker, rate-responsive	CT141	Electrodes/Leads and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	16/10/2020	16/10/2020
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	16/10/2020	16/10/2020
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery	16/10/2020	16/10/2020
					CT145	Cardiovascular devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	16/10/2020	16/10/2020
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	16/10/2020	16/10/2020
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery	16/10/2020	16/10/2020
					CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	16/10/2020	16/10/2020
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	16/10/2020	16/10/2020
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery	16/10/2020	16/10/2020
					CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	16/10/2020	16/10/2020
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	16/10/2020	16/10/2020
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery	16/10/2020	16/10/2020
					CT233	Surgical	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	16/10/2020	16/10/2020
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	16/10/2020	16/10/2020
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery	16/10/2020	16/10/2020
					CT2406	Implantable	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	16/10/2020	16/10/2020
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	16/10/2020	16/10/2020
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery	16/10/2020	16/10/2020
					CT2407	Active implantable	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	16/10/2020	16/10/2020
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	16/10/2020	16/10/2020
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery	16/10/2020	16/10/2020
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	16/10/2020	16/10/2020
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	16/10/2020	16/10/2020
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery	16/10/2020	16/10/2020
					CT2727	Active implant devices/ancillaries	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	16/10/2020	16/10/2020
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	16/10/2020	16/10/2020
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery	16/10/2020	16/10/2020
					CT2750	Implantable cardiac electrical stimulators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	16/10/2020	16/10/2020
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	16/10/2020	16/10/2020
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery	16/10/2020	16/10/2020
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	16/10/2020	16/10/2020
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	16/10/2020	16/10/2020
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery	16/10/2020	16/10/2020
					CT2985	Active	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	16/10/2020	16/10/2020
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	16/10/2020	16/10/2020
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery	16/10/2020	16/10/2020
					CT3001	Implantable metal	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	16/10/2020	16/10/2020
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	16/10/2020	16/10/2020
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery	16/10/2020	16/10/2020
					CT3002	Implantable synthetic polymer	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	16/10/2020	16/10/2020
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	16/10/2020	16/10/2020
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery	16/10/2020	16/10/2020
					CT3006	Implantable inorganic material	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	16/10/2020	16/10/2020
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	16/10/2020	16/10/2020
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery	16/10/2020	16/10/2020
					CT314	Electrically powered	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	16/10/2020	16/10/2020
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	16/10/2020	16/10/2020
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery	16/10/2020	16/10/2020
					CT321	Long-term surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	16/10/2020	16/10/2020
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	16/10/2020	16/10/2020
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery	16/10/2020	16/10/2020
					CT449	Generators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	16/10/2020	16/10/2020
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	16/10/2020	16/10/2020
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery	16/10/2020	16/10/2020
					CT981	Single use	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	16/10/2020	16/10/2020
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	16/10/2020	16/10/2020
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery	16/10/2020	16/10/2020
					CT983	Surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	16/10/2020	16/10/2020
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	16/10/2020	16/10/2020

							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery	16/10/2020	16/10/2020
10196	2020	All other reportable incidents	47267	Single-chamber implantable pacemaker, rate-responsive	CT141	Electrodes/Leads and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	01/12/2020	01/12/2020
							F19	Surgical Intervention	A070504	Premature Discharge of Battery	01/12/2020	01/12/2020
					CT145	Cardiovascular devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	01/12/2020	01/12/2020
							F19	Surgical Intervention	A070504	Premature Discharge of Battery	01/12/2020	01/12/2020
					CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	01/12/2020	01/12/2020
							F19	Surgical Intervention	A070504	Premature Discharge of Battery	01/12/2020	01/12/2020
					CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	01/12/2020	01/12/2020
							F19	Surgical Intervention	A070504	Premature Discharge of Battery	01/12/2020	01/12/2020
					CT233	Surgical	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	01/12/2020	01/12/2020
							F19	Surgical Intervention	A070504	Premature Discharge of Battery	01/12/2020	01/12/2020
					CT2406	Implantable	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	01/12/2020	01/12/2020
							F19	Surgical Intervention	A070504	Premature Discharge of Battery	01/12/2020	01/12/2020
					CT2407	Active implantable	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	01/12/2020	01/12/2020
							F19	Surgical Intervention	A070504	Premature Discharge of Battery	01/12/2020	01/12/2020
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	01/12/2020	01/12/2020
							F19	Surgical Intervention	A070504	Premature Discharge of Battery	01/12/2020	01/12/2020
					CT2727	Active implant devices/ancillaries	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	01/12/2020	01/12/2020
							F19	Surgical Intervention	A070504	Premature Discharge of Battery	01/12/2020	01/12/2020
					CT2750	Implantable cardiac electrical stimulators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	01/12/2020	01/12/2020
							F19	Surgical Intervention	A070504	Premature Discharge of Battery	01/12/2020	01/12/2020
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	01/12/2020	01/12/2020
							F19	Surgical Intervention	A070504	Premature Discharge of Battery	01/12/2020	01/12/2020
					CT2985	Active	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	01/12/2020	01/12/2020
							F19	Surgical Intervention	A070504	Premature Discharge of Battery	01/12/2020	01/12/2020
					CT3001	Implantable metal	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	01/12/2020	01/12/2020
							F19	Surgical Intervention	A070504	Premature Discharge of Battery	01/12/2020	01/12/2020
					CT3002	Implantable synthetic polymer	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	01/12/2020	01/12/2020
							F19	Surgical Intervention	A070504	Premature Discharge of Battery	01/12/2020	01/12/2020
					CT3006	Implantable inorganic material	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	01/12/2020	01/12/2020
							F19	Surgical Intervention	A070504	Premature Discharge of Battery	01/12/2020	01/12/2020
					CT314	Electrically powered	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	01/12/2020	01/12/2020
							F19	Surgical Intervention	A070504	Premature Discharge of Battery	01/12/2020	01/12/2020
					CT321	Long-term surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	01/12/2020	01/12/2020
							F19	Surgical Intervention	A070504	Premature Discharge of Battery	01/12/2020	01/12/2020
					CT449	Generators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	01/12/2020	01/12/2020
							F19	Surgical Intervention	A070504	Premature Discharge of Battery	01/12/2020	01/12/2020
					CT981	Single use	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	01/12/2020	01/12/2020
							F19	Surgical Intervention	A070504	Premature Discharge of Battery	01/12/2020	01/12/2020
					CT983	Surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	01/12/2020	01/12/2020
							F19	Surgical Intervention	A070504	Premature Discharge of Battery	01/12/2020	01/12/2020
10197	2020	All other reportable incidents	37265	Dual-chamber implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A0712	Pacing Problem		
									A0909	Unexpected Therapeutic Results		
							F24	Insufficient Health Impact Information	A0712	Pacing Problem		
									A0909	Unexpected Therapeutic Results		
10198	2020	All other reportable incidents	47270	Cardiac resynchronization therapy implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A0908	Incorrect, Inadequate or Imprecise Result or Readings	23/09/2020	23/09/2020
							F14	Prolonged Episode of Care	A0908	Incorrect, Inadequate or Imprecise Result or Readings	23/09/2020	23/09/2020
							F1905	Device Revision or Replacement	A0908	Incorrect, Inadequate or Imprecise Result or Readings	23/09/2020	23/09/2020
							F1908	Prolonged Surgery	A0908	Incorrect, Inadequate or Imprecise Result or Readings	23/09/2020	23/09/2020
10199	2020	All other reportable incidents	36241	Pacemaker/defibrillator lead	CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A051201	Device Dislodged or Dislocated		
									A0712	Pacing Problem		
							F26	No Health Consequences or Impact	A051201	Device Dislodged or Dislocated		
									A0712	Pacing Problem		
10200	2020	All other reportable incidents	47265	Dual-chamber implantable pacemaker, rate-responsive	CT141	Electrodes/Leads and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
					CT145	Cardiovascular devices	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
					CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
					CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
					CT233	Surgical	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
					CT2406	Implantable	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
					CT2407	Active implantable	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		

					CT2727	Active implant devices/ancillaries	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
					CT2750	Implantable cardiac electrical stimulators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
					CT2985	Active	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
					CT3001	Implantable metal	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
					CT3002	Implantable synthetic polymer	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
					CT3006	Implantable inorganic material	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
					CT314	Electrically powered	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
					CT321	Long-term surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
					CT449	Generators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
					CT981	Single use	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
					CT983	Surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
10201	2020	All other reportable incidents	47270	Cardiac resynchronization therapy implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	23/10/2020	23/10/2020
							F1901	Additional Surgery	A070504	Premature Discharge of Battery	23/10/2020	23/10/2020
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	23/10/2020	23/10/2020
10202	2020	Death	47270	Cardiac resynchronization therapy implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E0601	Arrhythmia	A070910	Undersensing	16/10/2020	16/10/2020
									A070912	Incorrect Interpretation of Signal	16/10/2020	16/10/2020
									A071301	Failure to Deliver Shock/Stimulation	16/10/2020	16/10/2020
							F02	Death	A070910	Undersensing	16/10/2020	16/10/2020
									A070912	Incorrect Interpretation of Signal	16/10/2020	16/10/2020
									A071301	Failure to Deliver Shock/Stimulation	16/10/2020	16/10/2020
10203	2020	All other reportable incidents	47270	Cardiac resynchronization therapy implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	10/06/2020	10/06/2020
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	10/06/2020	10/06/2020
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery	10/06/2020	10/06/2020
10204	2020	All other reportable incidents	36241	Pacemaker/defibrillator lead	CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A072201	High Impedance		
							F26	No Health Consequences or Impact	A072201	High Impedance		
10205	2020	All other reportable incidents	35852	Single-chamber implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E2401	Insufficient Information	A072201	High Impedance		
							E2403	No Clinical Signs, Symptoms or Conditions	A072201	High Impedance		
							F26	No Health Consequences or Impact	A072201	High Impedance		
10206	2020	All other reportable incidents	47265	Dual-chamber implantable pacemaker, rate-responsive	CT141	Electrodes/Leads and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT145	Cardiovascular devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT233	Surgical	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT2406	Implantable	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT2407	Active implantable	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT2727	Active implant devices/ancillaries	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT2750	Implantable cardiac electrical stimulators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT2985	Active	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		

							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT3001	Implantable metal	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT3002	Implantable synthetic polymer	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT3006	Implantable inorganic material	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT314	Electrically powered	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT321	Long-term surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT449	Generators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT981	Single use	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT983	Surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
10207	2020	All other reportable incidents	47263	Cardiac resynchronization therapy implantable pacemaker	CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A072201	High Impedance		
									A072202	Low Impedance		
							F24	Insufficient Health Impact Information	A072201	High Impedance		
									A072202	Low Impedance		
					CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A072201	High Impedance		
									A072202	Low Impedance		
							F24	Insufficient Health Impact Information	A072201	High Impedance		
									A072202	Low Impedance		
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A072201	High Impedance		
									A072202	Low Impedance		
							F24	Insufficient Health Impact Information	A072201	High Impedance		
									A072202	Low Impedance		
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A072201	High Impedance		
									A072202	Low Impedance		
							F24	Insufficient Health Impact Information	A072201	High Impedance		
									A072202	Low Impedance		
10208	2020	All other reportable incidents	47265	Dual-chamber implantable pacemaker, rate-responsive	CT141	Electrodes/Leads and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT145	Cardiovascular devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT233	Surgical	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT2406	Implantable	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT2407	Active implantable	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT2727	Active implant devices/ancillaries	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT2750	Implantable cardiac electrical stimulators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT2985	Active	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT3001	Implantable metal	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT3002	Implantable synthetic polymer	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT3006	Implantable inorganic material	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT314	Electrically powered	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT321	Long-term surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT449	Generators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT981	Single use	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT983	Surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
10209	2020	All other reportable incidents	35852	Single-chamber implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E2104	Electric Shock	A071302	Inappropriate/Inadequate Shock/Stimulation	02/11/2020	02/11/2020
							F08	Hospitalization or Prolonged Hospitalization	A071302	Inappropriate/Inadequate Shock/Stimulation	02/11/2020	02/11/2020
							F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A071302	Inappropriate/Inadequate Shock/Stimulation	02/11/2020	02/11/2020
							F19	Surgical Intervention	A071302	Inappropriate/Inadequate Shock/Stimulation	02/11/2020	02/11/2020
							F1905	Device Revision or Replacement	A071302	Inappropriate/Inadequate Shock/Stimulation	02/11/2020	02/11/2020
10210	2020	All other reportable incidents	35852	Single-chamber implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E2104	Electric Shock	A071302	Inappropriate/Inadequate Shock/Stimulation		
							F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A071302	Inappropriate/Inadequate Shock/Stimulation		
							F26	No Health Consequences or Impact	A071302	Inappropriate/Inadequate Shock/Stimulation		
10211	2020	All other reportable incidents	17882	Physiologic-monitoring defibrillation system	CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A07	Electrical /Electronic Property Problem		
							F26	No Health Consequences or Impact	A07	Electrical /Electronic Property Problem		
10212	2020	All other reportable incidents	17882	Physiologic-monitoring defibrillation system	CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070802	Intermittent Loss of Power		
							F26	No Health Consequences or Impact	A070802	Intermittent Loss of Power		
10213	2020	All other reportable incidents	17882	Physiologic-monitoring defibrillation system	CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070802	Intermittent Loss of Power		
							F26	No Health Consequences or Impact	A070802	Intermittent Loss of Power		
10214	2020	All other reportable incidents	47265	Dual-chamber implantable pacemaker, rate-	CT141	Electrodes/Leads and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A071205	Pacemaker Found in Back-Up Mode	14/10/2020	14/10/2020

				responsive			F1905	Device Revision or Replacement	A071205	Pacemaker Found in Back-Up Mode	14/10/2020	14/10/2020
							F1908	Prolonged Surgery	A071205	Pacemaker Found in Back-Up Mode	14/10/2020	14/10/2020
					CT145	Cardiovascular devices	E2403	No Clinical Signs, Symptoms or Conditions	A071205	Pacemaker Found in Back-Up Mode	14/10/2020	14/10/2020
							F1905	Device Revision or Replacement	A071205	Pacemaker Found in Back-Up Mode	14/10/2020	14/10/2020
							F1908	Prolonged Surgery	A071205	Pacemaker Found in Back-Up Mode	14/10/2020	14/10/2020
					CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A071205	Pacemaker Found in Back-Up Mode	14/10/2020	14/10/2020
							F1905	Device Revision or Replacement	A071205	Pacemaker Found in Back-Up Mode	14/10/2020	14/10/2020
							F1908	Prolonged Surgery	A071205	Pacemaker Found in Back-Up Mode	14/10/2020	14/10/2020
					CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A071205	Pacemaker Found in Back-Up Mode	14/10/2020	14/10/2020
							F1905	Device Revision or Replacement	A071205	Pacemaker Found in Back-Up Mode	14/10/2020	14/10/2020
							F1908	Prolonged Surgery	A071205	Pacemaker Found in Back-Up Mode	14/10/2020	14/10/2020
					CT233	Surgical	E2403	No Clinical Signs, Symptoms or Conditions	A071205	Pacemaker Found in Back-Up Mode	14/10/2020	14/10/2020
							F1905	Device Revision or Replacement	A071205	Pacemaker Found in Back-Up Mode	14/10/2020	14/10/2020
							F1908	Prolonged Surgery	A071205	Pacemaker Found in Back-Up Mode	14/10/2020	14/10/2020
					CT2406	Implantable	E2403	No Clinical Signs, Symptoms or Conditions	A071205	Pacemaker Found in Back-Up Mode	14/10/2020	14/10/2020
							F1905	Device Revision or Replacement	A071205	Pacemaker Found in Back-Up Mode	14/10/2020	14/10/2020
							F1908	Prolonged Surgery	A071205	Pacemaker Found in Back-Up Mode	14/10/2020	14/10/2020
					CT2407	Active implantable	E2403	No Clinical Signs, Symptoms or Conditions	A071205	Pacemaker Found in Back-Up Mode	14/10/2020	14/10/2020
							F1905	Device Revision or Replacement	A071205	Pacemaker Found in Back-Up Mode	14/10/2020	14/10/2020
							F1908	Prolonged Surgery	A071205	Pacemaker Found in Back-Up Mode	14/10/2020	14/10/2020
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A071205	Pacemaker Found in Back-Up Mode	14/10/2020	14/10/2020
							F1905	Device Revision or Replacement	A071205	Pacemaker Found in Back-Up Mode	14/10/2020	14/10/2020
							F1908	Prolonged Surgery	A071205	Pacemaker Found in Back-Up Mode	14/10/2020	14/10/2020
					CT2727	Active implant devices/ancillaries	E2403	No Clinical Signs, Symptoms or Conditions	A071205	Pacemaker Found in Back-Up Mode	14/10/2020	14/10/2020
							F1905	Device Revision or Replacement	A071205	Pacemaker Found in Back-Up Mode	14/10/2020	14/10/2020
							F1908	Prolonged Surgery	A071205	Pacemaker Found in Back-Up Mode	14/10/2020	14/10/2020
					CT2750	Implantable cardiac electrical stimulators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A071205	Pacemaker Found in Back-Up Mode	14/10/2020	14/10/2020
							F1905	Device Revision or Replacement	A071205	Pacemaker Found in Back-Up Mode	14/10/2020	14/10/2020
							F1908	Prolonged Surgery	A071205	Pacemaker Found in Back-Up Mode	14/10/2020	14/10/2020
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A071205	Pacemaker Found in Back-Up Mode	14/10/2020	14/10/2020
							F1905	Device Revision or Replacement	A071205	Pacemaker Found in Back-Up Mode	14/10/2020	14/10/2020
							F1908	Prolonged Surgery	A071205	Pacemaker Found in Back-Up Mode	14/10/2020	14/10/2020
					CT2985	Active	E2403	No Clinical Signs, Symptoms or Conditions	A071205	Pacemaker Found in Back-Up Mode	14/10/2020	14/10/2020
							F1905	Device Revision or Replacement	A071205	Pacemaker Found in Back-Up Mode	14/10/2020	14/10/2020
							F1908	Prolonged Surgery	A071205	Pacemaker Found in Back-Up Mode	14/10/2020	14/10/2020
					CT3001	Implantable metal	E2403	No Clinical Signs, Symptoms or Conditions	A071205	Pacemaker Found in Back-Up Mode	14/10/2020	14/10/2020
							F1905	Device Revision or Replacement	A071205	Pacemaker Found in Back-Up Mode	14/10/2020	14/10/2020
							F1908	Prolonged Surgery	A071205	Pacemaker Found in Back-Up Mode	14/10/2020	14/10/2020
					CT3002	Implantable synthetic polymer	E2403	No Clinical Signs, Symptoms or Conditions	A071205	Pacemaker Found in Back-Up Mode	14/10/2020	14/10/2020
							F1905	Device Revision or Replacement	A071205	Pacemaker Found in Back-Up Mode	14/10/2020	14/10/2020
							F1908	Prolonged Surgery	A071205	Pacemaker Found in Back-Up Mode	14/10/2020	14/10/2020
					CT3006	Implantable inorganic material	E2403	No Clinical Signs, Symptoms or Conditions	A071205	Pacemaker Found in Back-Up Mode	14/10/2020	14/10/2020
							F1905	Device Revision or Replacement	A071205	Pacemaker Found in Back-Up Mode	14/10/2020	14/10/2020
							F1908	Prolonged Surgery	A071205	Pacemaker Found in Back-Up Mode	14/10/2020	14/10/2020
					CT314	Electrically powered	E2403	No Clinical Signs, Symptoms or Conditions	A071205	Pacemaker Found in Back-Up Mode	14/10/2020	14/10/2020
							F1905	Device Revision or Replacement	A071205	Pacemaker Found in Back-Up Mode	14/10/2020	14/10/2020
							F1908	Prolonged Surgery	A071205	Pacemaker Found in Back-Up Mode	14/10/2020	14/10/2020
					CT321	Long-term surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A071205	Pacemaker Found in Back-Up Mode	14/10/2020	14/10/2020
							F1905	Device Revision or Replacement	A071205	Pacemaker Found in Back-Up Mode	14/10/2020	14/10/2020
							F1908	Prolonged Surgery	A071205	Pacemaker Found in Back-Up Mode	14/10/2020	14/10/2020
					CT449	Generators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A071205	Pacemaker Found in Back-Up Mode	14/10/2020	14/10/2020
							F1905	Device Revision or Replacement	A071205	Pacemaker Found in Back-Up Mode	14/10/2020	14/10/2020
							F1908	Prolonged Surgery	A071205	Pacemaker Found in Back-Up Mode	14/10/2020	14/10/2020
					CT981	Single use	E2403	No Clinical Signs, Symptoms or Conditions	A071205	Pacemaker Found in Back-Up Mode	14/10/2020	14/10/2020
							F1905	Device Revision or Replacement	A071205	Pacemaker Found in Back-Up Mode	14/10/2020	14/10/2020
							F1908	Prolonged Surgery	A071205	Pacemaker Found in Back-Up Mode	14/10/2020	14/10/2020
					CT983	Surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A071205	Pacemaker Found in Back-Up Mode	14/10/2020	14/10/2020
							F1905	Device Revision or Replacement	A071205	Pacemaker Found in Back-Up Mode	14/10/2020	14/10/2020
							F1908	Prolonged Surgery	A071205	Pacemaker Found in Back-Up Mode	14/10/2020	14/10/2020
10215	2020	All other reportable incidents	35223	Endocardial/interventricular septal pacing lead	CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A0401	Break		
									A070909	Oversensing		
							F1901	Additional Surgery	A0401	Break		
									A070909	Oversensing		
							F1905	Device Revision or Replacement	A0401	Break		
									A070909	Oversensing		
10216	2020	All other reportable incidents	47270	Cardiac resynchronization therapy implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A070909	Oversensing		
									A0712	Pacing Problem		
							F26	No Health Consequences or Impact	A070909	Oversensing		
									A0712	Pacing Problem		
10217	2020	All other reportable incidents	17882	Physiologic-monitoring defibrillation system	CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A0719	Unexpected Shutdown		
							F26	No Health Consequences or Impact	A0719	Unexpected Shutdown		
10218	2020	All other reportable incidents	37265	Dual-chamber implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A070504	Premature Discharge of Battery		
							F15	Recognized Device or Procedural Complication	A070504	Premature Discharge of Battery		
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery		
10219	2020	All other reportable incidents	37265	Dual-chamber implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A04	Material Integrity Problem	25/09/2020	25/09/2020
									A072201	High Impedance	25/09/2020	25/09/2020
							F19	Surgical Intervention	A04	Material Integrity Problem	25/09/2020	25/09/2020
									A072201	High Impedance	25/09/2020	25/09/2020
							F1905	Device Revision or Replacement	A04	Material Integrity Problem	25/09/2020	25/09/2020
									A072201	High Impedance	25/09/2020	25/09/2020
10220	2020	All other reportable incidents	47263	Cardiac resynchronization therapy implantable pacemaker	CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A090805	Incorrect Measurement	19/10/2020	19/10/2020
							F1901	Additional Surgery	A090805	Incorrect Measurement	19/10/2020	19/10/2020
							F1905	Device Revision or Replacement	A090805	Incorrect Measurement	19/10/2020	19/10/2020
					CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A090805	Incorrect Measurement	19/10/2020	19/10/2020
							F1901	Additional Surgery	A090805	Incorrect Measurement	19/10/2020	19/10/2020
							F1905	Device Revision or Replacement	A090805	Incorrect Measurement	19/10/2020	19/10/2020
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A090805	Incorrect Measurement	19/10/2020	19/10/2020

						F1901	Additional Surgery	A090805	Incorrect Measurement	19/10/2020	19/10/2020
						F1905	Device Revision or Replacement	A090805	Incorrect Measurement	19/10/2020	19/10/2020
				CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A090805	Incorrect Measurement	19/10/2020	19/10/2020
						F1901	Additional Surgery	A090805	Incorrect Measurement	19/10/2020	19/10/2020
						F1905	Device Revision or Replacement	A090805	Incorrect Measurement	19/10/2020	19/10/2020
10221	2020	All other reportable incidents	47265	Dual-chamber implantable pacemaker, rate-responsive	CT141	Electrodes/Leads and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	26/07/2020	26/07/2020
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	26/07/2020	26/07/2020
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery	26/07/2020	26/07/2020
					CT145	Cardiovascular devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	26/07/2020	26/07/2020
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	26/07/2020	26/07/2020
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery	26/07/2020	26/07/2020
					CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	26/07/2020	26/07/2020
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	26/07/2020	26/07/2020
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery	26/07/2020	26/07/2020
					CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	26/07/2020	26/07/2020
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	26/07/2020	26/07/2020
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery	26/07/2020	26/07/2020
					CT233	Surgical	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	26/07/2020	26/07/2020
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	26/07/2020	26/07/2020
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery	26/07/2020	26/07/2020
					CT2406	Implantable	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	26/07/2020	26/07/2020
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	26/07/2020	26/07/2020
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery	26/07/2020	26/07/2020
					CT2407	Active implantable	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	26/07/2020	26/07/2020
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	26/07/2020	26/07/2020
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery	26/07/2020	26/07/2020
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	26/07/2020	26/07/2020
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	26/07/2020	26/07/2020
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery	26/07/2020	26/07/2020
					CT2727	Active implant devices/ancillaries	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	26/07/2020	26/07/2020
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	26/07/2020	26/07/2020
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery	26/07/2020	26/07/2020
					CT2750	Implantable cardiac electrical stimulators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	26/07/2020	26/07/2020
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	26/07/2020	26/07/2020
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery	26/07/2020	26/07/2020
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	26/07/2020	26/07/2020
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	26/07/2020	26/07/2020
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery	26/07/2020	26/07/2020
					CT2985	Active	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	26/07/2020	26/07/2020
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	26/07/2020	26/07/2020
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery	26/07/2020	26/07/2020
					CT3001	Implantable metal	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	26/07/2020	26/07/2020
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	26/07/2020	26/07/2020
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery	26/07/2020	26/07/2020
					CT3002	Implantable synthetic polymer	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	26/07/2020	26/07/2020
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	26/07/2020	26/07/2020
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery	26/07/2020	26/07/2020
					CT3006	Implantable inorganic material	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	26/07/2020	26/07/2020
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	26/07/2020	26/07/2020
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery	26/07/2020	26/07/2020
					CT314	Electrically powered	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	26/07/2020	26/07/2020
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	26/07/2020	26/07/2020
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery	26/07/2020	26/07/2020
					CT321	Long-term surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	26/07/2020	26/07/2020
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	26/07/2020	26/07/2020
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery	26/07/2020	26/07/2020
					CT449	Generators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	26/07/2020	26/07/2020
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	26/07/2020	26/07/2020
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery	26/07/2020	26/07/2020
					CT981	Single use	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	26/07/2020	26/07/2020
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	26/07/2020	26/07/2020
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery	26/07/2020	26/07/2020
					CT983	Surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	26/07/2020	26/07/2020
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	26/07/2020	26/07/2020
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery	26/07/2020	26/07/2020
10222	2020	All other reportable incidents	47263	Cardiac resynchronization therapy implantable pacemaker	CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A072201	High Impedance		
							F26	No Health Consequences or Impact	A072201	High Impedance		
					CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A072201	High Impedance		
							F26	No Health Consequences or Impact	A072201	High Impedance		
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A072201	High Impedance		
							F26	No Health Consequences or Impact	A072201	High Impedance		
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A072201	High Impedance		
							F26	No Health Consequences or Impact	A072201	High Impedance		
10223	2020	All other reportable incidents	35852	Single-chamber implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	30/11/2020	30/11/2020
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	30/11/2020	30/11/2020
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	30/11/2020	30/11/2020
10224	2020	All other reportable incidents	47265	Dual-chamber implantable pacemaker, rate-responsive	CT141	Electrodes/Leads and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery		
							F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A070504	Premature Discharge of Battery		
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery		
					CT145	Cardiovascular devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery		
							F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A070504	Premature Discharge of Battery		
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery		
					CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery		
							F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A070504	Premature Discharge of Battery		
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery		

					CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery		
							F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A070504	Premature Discharge of Battery		
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery		
					CT233	Surgical	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery		
							F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A070504	Premature Discharge of Battery		
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery		
					CT2406	Implantable	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery		
							F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A070504	Premature Discharge of Battery		
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery		
					CT2407	Active implantable	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery		
							F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A070504	Premature Discharge of Battery		
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery		
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery		
							F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A070504	Premature Discharge of Battery		
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery		
					CT2727	Active implant devices/ancillaries	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery		
							F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A070504	Premature Discharge of Battery		
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery		
					CT2750	Implantable cardiac electrical stimulators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery		
							F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A070504	Premature Discharge of Battery		
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery		
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery		
							F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A070504	Premature Discharge of Battery		
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery		
					CT2985	Active	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery		
							F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A070504	Premature Discharge of Battery		
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery		
					CT3001	Implantable metal	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery		
							F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A070504	Premature Discharge of Battery		
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery		
					CT3002	Implantable synthetic polymer	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery		
							F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A070504	Premature Discharge of Battery		
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery		
					CT3006	Implantable inorganic material	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery		
							F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A070504	Premature Discharge of Battery		
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery		
					CT314	Electrically powered	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery		
							F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A070504	Premature Discharge of Battery		
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery		
					CT321	Long-term surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery		
							F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A070504	Premature Discharge of Battery		
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery		
					CT449	Generators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery		
							F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A070504	Premature Discharge of Battery		
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery		
					CT981	Single use	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery		
							F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A070504	Premature Discharge of Battery		
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery		
					CT983	Surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery		
							F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A070504	Premature Discharge of Battery		
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery		
10225	2020	All other reportable incidents	47265	Dual-chamber implantable pacemaker, rate-responsive	CT141	Electrodes/Leads and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem	19/11/2021	19/11/2021
									A070504	Premature Discharge of Battery	19/11/2021	19/11/2021
							F08	Hospitalization or Prolonged Hospitalization	A0705	Battery Problem	19/11/2021	19/11/2021
									A070504	Premature Discharge of Battery	19/11/2021	19/11/2021
							F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A0705	Battery Problem	19/11/2021	19/11/2021
									A070504	Premature Discharge of Battery	19/11/2021	19/11/2021
							F1903	Device Explantation	A0705	Battery Problem	19/11/2021	19/11/2021
									A070504	Premature Discharge of Battery	19/11/2021	19/11/2021
					CT145	Cardiovascular devices	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem	19/11/2021	19/11/2021
									A070504	Premature Discharge of Battery	19/11/2021	19/11/2021
							F08	Hospitalization or Prolonged Hospitalization	A0705	Battery Problem	19/11/2021	19/11/2021
									A070504	Premature Discharge of Battery	19/11/2021	19/11/2021
							F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A0705	Battery Problem	19/11/2021	19/11/2021
									A070504	Premature Discharge of Battery	19/11/2021	19/11/2021
							F1903	Device Explantation	A0705	Battery Problem	19/11/2021	19/11/2021
									A070504	Premature Discharge of Battery	19/11/2021	19/11/2021
					CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem	19/11/2021	19/11/2021
									A070504	Premature Discharge of Battery	19/11/2021	19/11/2021
							F08	Hospitalization or Prolonged Hospitalization	A0705	Battery Problem	19/11/2021	19/11/2021
									A070504	Premature Discharge of Battery	19/11/2021	19/11/2021

							F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A0705	Battery Problem	19/11/2021	19/11/2021
									A070504	Premature Discharge of Battery	19/11/2021	19/11/2021
							F1903	Device Explantation	A0705	Battery Problem	19/11/2021	19/11/2021
									A070504	Premature Discharge of Battery	19/11/2021	19/11/2021
					CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem	19/11/2021	19/11/2021
									A070504	Premature Discharge of Battery	19/11/2021	19/11/2021
							F08	Hospitalization or Prolonged Hospitalization	A0705	Battery Problem	19/11/2021	19/11/2021
									A070504	Premature Discharge of Battery	19/11/2021	19/11/2021
							F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A0705	Battery Problem	19/11/2021	19/11/2021
									A070504	Premature Discharge of Battery	19/11/2021	19/11/2021
							F1903	Device Explantation	A0705	Battery Problem	19/11/2021	19/11/2021
									A070504	Premature Discharge of Battery	19/11/2021	19/11/2021
					CT233	Surgical	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem	19/11/2021	19/11/2021
									A070504	Premature Discharge of Battery	19/11/2021	19/11/2021
							F08	Hospitalization or Prolonged Hospitalization	A0705	Battery Problem	19/11/2021	19/11/2021
									A070504	Premature Discharge of Battery	19/11/2021	19/11/2021
							F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A0705	Battery Problem	19/11/2021	19/11/2021
									A070504	Premature Discharge of Battery	19/11/2021	19/11/2021
							F1903	Device Explantation	A0705	Battery Problem	19/11/2021	19/11/2021
									A070504	Premature Discharge of Battery	19/11/2021	19/11/2021
					CT2406	Implantable	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem	19/11/2021	19/11/2021
									A070504	Premature Discharge of Battery	19/11/2021	19/11/2021
							F08	Hospitalization or Prolonged Hospitalization	A0705	Battery Problem	19/11/2021	19/11/2021
									A070504	Premature Discharge of Battery	19/11/2021	19/11/2021
							F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A0705	Battery Problem	19/11/2021	19/11/2021
									A070504	Premature Discharge of Battery	19/11/2021	19/11/2021
							F1903	Device Explantation	A0705	Battery Problem	19/11/2021	19/11/2021
									A070504	Premature Discharge of Battery	19/11/2021	19/11/2021
					CT2407	Active implantable	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem	19/11/2021	19/11/2021
									A070504	Premature Discharge of Battery	19/11/2021	19/11/2021
							F08	Hospitalization or Prolonged Hospitalization	A0705	Battery Problem	19/11/2021	19/11/2021
									A070504	Premature Discharge of Battery	19/11/2021	19/11/2021
							F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A0705	Battery Problem	19/11/2021	19/11/2021
									A070504	Premature Discharge of Battery	19/11/2021	19/11/2021
							F1903	Device Explantation	A0705	Battery Problem	19/11/2021	19/11/2021
									A070504	Premature Discharge of Battery	19/11/2021	19/11/2021
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem	19/11/2021	19/11/2021
									A070504	Premature Discharge of Battery	19/11/2021	19/11/2021
							F08	Hospitalization or Prolonged Hospitalization	A0705	Battery Problem	19/11/2021	19/11/2021
									A070504	Premature Discharge of Battery	19/11/2021	19/11/2021
							F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A0705	Battery Problem	19/11/2021	19/11/2021
									A070504	Premature Discharge of Battery	19/11/2021	19/11/2021
							F1903	Device Explantation	A0705	Battery Problem	19/11/2021	19/11/2021
									A070504	Premature Discharge of Battery	19/11/2021	19/11/2021
					CT2727	Active implant devices/ancillaries	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem	19/11/2021	19/11/2021
									A070504	Premature Discharge of Battery	19/11/2021	19/11/2021
							F08	Hospitalization or Prolonged Hospitalization	A0705	Battery Problem	19/11/2021	19/11/2021
									A070504	Premature Discharge of Battery	19/11/2021	19/11/2021
							F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A0705	Battery Problem	19/11/2021	19/11/2021
									A070504	Premature Discharge of Battery	19/11/2021	19/11/2021
							F1903	Device Explantation	A0705	Battery Problem	19/11/2021	19/11/2021
									A070504	Premature Discharge of Battery	19/11/2021	19/11/2021
					CT2750	Implantable cardiac electrical stimulators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem	19/11/2021	19/11/2021
									A070504	Premature Discharge of Battery	19/11/2021	19/11/2021
							F08	Hospitalization or Prolonged Hospitalization	A0705	Battery Problem	19/11/2021	19/11/2021
									A070504	Premature Discharge of Battery	19/11/2021	19/11/2021
							F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A0705	Battery Problem	19/11/2021	19/11/2021
									A070504	Premature Discharge of Battery	19/11/2021	19/11/2021
							F1903	Device Explantation	A0705	Battery Problem	19/11/2021	19/11/2021
									A070504	Premature Discharge of Battery	19/11/2021	19/11/2021
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem	19/11/2021	19/11/2021
									A070504	Premature Discharge of Battery	19/11/2021	19/11/2021
							F08	Hospitalization or Prolonged Hospitalization	A0705	Battery Problem	19/11/2021	19/11/2021
									A070504	Premature Discharge of Battery	19/11/2021	19/11/2021
							F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A0705	Battery Problem	19/11/2021	19/11/2021
									A070504	Premature Discharge of Battery	19/11/2021	19/11/2021
							F1903	Device Explantation	A0705	Battery Problem	19/11/2021	19/11/2021
									A070504	Premature Discharge of Battery	19/11/2021	19/11/2021
					CT2985	Active	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem	19/11/2021	19/11/2021
									A070504	Premature Discharge of Battery	19/11/2021	19/11/2021
							F08	Hospitalization or Prolonged Hospitalization	A0705	Battery Problem	19/11/2021	19/11/2021
									A070504	Premature Discharge of Battery	19/11/2021	19/11/2021
							F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A0705	Battery Problem	19/11/2021	19/11/2021
									A070504	Premature Discharge of Battery	19/11/2021	19/11/2021
							F1903	Device Explantation	A0705	Battery Problem	19/11/2021	19/11/2021
									A070504	Premature Discharge of Battery	19/11/2021	19/11/2021
					CT3001	Implantable metal	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem	19/11/2021	19/11/2021
									A070504	Premature Discharge of Battery	19/11/2021	19/11/2021
							F08	Hospitalization or Prolonged Hospitalization	A0705	Battery Problem	19/11/2021	19/11/2021
									A070504	Premature Discharge of Battery	19/11/2021	19/11/2021
							F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A0705	Battery Problem	19/11/2021	19/11/2021
									A070504	Premature Discharge of Battery	19/11/2021	19/11/2021
							F1903	Device Explantation	A0705	Battery Problem	19/11/2021	19/11/2021
									A070504	Premature Discharge of Battery	19/11/2021	19/11/2021
					CT3002	Implantable synthetic polymer	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem	19/11/2021	19/11/2021
									A070504	Premature Discharge of Battery	19/11/2021	19/11/2021
							F08	Hospitalization or Prolonged Hospitalization	A0705	Battery Problem	19/11/2021	19/11/2021
									A070504	Premature Discharge of Battery	19/11/2021	19/11/2021

							F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A0705	Battery Problem	19/11/2021	19/11/2021
									A070504	Premature Discharge of Battery	19/11/2021	19/11/2021
							F1903	Device Explantation	A0705	Battery Problem	19/11/2021	19/11/2021
									A070504	Premature Discharge of Battery	19/11/2021	19/11/2021
					CT3006	Implantable inorganic material	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem	19/11/2021	19/11/2021
									A070504	Premature Discharge of Battery	19/11/2021	19/11/2021
							F08	Hospitalization or Prolonged Hospitalization	A0705	Battery Problem	19/11/2021	19/11/2021
									A070504	Premature Discharge of Battery	19/11/2021	19/11/2021
							F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A0705	Battery Problem	19/11/2021	19/11/2021
									A070504	Premature Discharge of Battery	19/11/2021	19/11/2021
							F1903	Device Explantation	A0705	Battery Problem	19/11/2021	19/11/2021
									A070504	Premature Discharge of Battery	19/11/2021	19/11/2021
					CT314	Electrically powered	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem	19/11/2021	19/11/2021
									A070504	Premature Discharge of Battery	19/11/2021	19/11/2021
							F08	Hospitalization or Prolonged Hospitalization	A0705	Battery Problem	19/11/2021	19/11/2021
									A070504	Premature Discharge of Battery	19/11/2021	19/11/2021
							F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A0705	Battery Problem	19/11/2021	19/11/2021
									A070504	Premature Discharge of Battery	19/11/2021	19/11/2021
							F1903	Device Explantation	A0705	Battery Problem	19/11/2021	19/11/2021
									A070504	Premature Discharge of Battery	19/11/2021	19/11/2021
					CT321	Long-term surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem	19/11/2021	19/11/2021
									A070504	Premature Discharge of Battery	19/11/2021	19/11/2021
							F08	Hospitalization or Prolonged Hospitalization	A0705	Battery Problem	19/11/2021	19/11/2021
									A070504	Premature Discharge of Battery	19/11/2021	19/11/2021
							F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A0705	Battery Problem	19/11/2021	19/11/2021
									A070504	Premature Discharge of Battery	19/11/2021	19/11/2021
							F1903	Device Explantation	A0705	Battery Problem	19/11/2021	19/11/2021
									A070504	Premature Discharge of Battery	19/11/2021	19/11/2021
					CT449	Generators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem	19/11/2021	19/11/2021
									A070504	Premature Discharge of Battery	19/11/2021	19/11/2021
							F08	Hospitalization or Prolonged Hospitalization	A0705	Battery Problem	19/11/2021	19/11/2021
									A070504	Premature Discharge of Battery	19/11/2021	19/11/2021
							F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A0705	Battery Problem	19/11/2021	19/11/2021
									A070504	Premature Discharge of Battery	19/11/2021	19/11/2021
							F1903	Device Explantation	A0705	Battery Problem	19/11/2021	19/11/2021
									A070504	Premature Discharge of Battery	19/11/2021	19/11/2021
					CT981	Single use	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem	19/11/2021	19/11/2021
									A070504	Premature Discharge of Battery	19/11/2021	19/11/2021
							F08	Hospitalization or Prolonged Hospitalization	A0705	Battery Problem	19/11/2021	19/11/2021
									A070504	Premature Discharge of Battery	19/11/2021	19/11/2021
							F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A0705	Battery Problem	19/11/2021	19/11/2021
									A070504	Premature Discharge of Battery	19/11/2021	19/11/2021
							F1903	Device Explantation	A0705	Battery Problem	19/11/2021	19/11/2021
									A070504	Premature Discharge of Battery	19/11/2021	19/11/2021
					CT983	Surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem	19/11/2021	19/11/2021
									A070504	Premature Discharge of Battery	19/11/2021	19/11/2021
							F08	Hospitalization or Prolonged Hospitalization	A0705	Battery Problem	19/11/2021	19/11/2021
									A070504	Premature Discharge of Battery	19/11/2021	19/11/2021
							F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A0705	Battery Problem	19/11/2021	19/11/2021
									A070504	Premature Discharge of Battery	19/11/2021	19/11/2021
							F1903	Device Explantation	A0705	Battery Problem	19/11/2021	19/11/2021
									A070504	Premature Discharge of Battery	19/11/2021	19/11/2021
10226	2020	All other reportable incidents	35852	Single-chamber implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	27/11/2020	27/11/2020
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	27/11/2020	27/11/2020
							F12	Serious Injury/ Illness/ Impairment	A070504	Premature Discharge of Battery	27/11/2020	27/11/2020
							F19	Surgical Intervention	A070504	Premature Discharge of Battery	27/11/2020	27/11/2020
10227	2020	All other reportable incidents	37265	Dual-chamber implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
10228	2020	All other reportable incidents	36241	Pacemaker/defibrillator lead	CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070909	Oversensing		
									A090801	Signal Artifact/Noise		
							F08	Hospitalization or Prolonged Hospitalization	A070909	Oversensing		
									A090801	Signal Artifact/Noise		
							F14	Prolonged Episode of Care	A070909	Oversensing		
									A090801	Signal Artifact/Noise		
10229	2020	All other reportable incidents	37265	Dual-chamber implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A090801	Signal Artifact/Noise		
							F26	No Health Consequences or Impact	A090801	Signal Artifact/Noise		
10230	2020	All other reportable incidents	47265	Dual-chamber implantable pacemaker, rate-responsive	CT141	Electrodes/Leads and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A090804	False Positive Result		
							F26	No Health Consequences or Impact	A090804	False Positive Result		
					CT145	Cardiovascular devices	E2403	No Clinical Signs, Symptoms or Conditions	A090804	False Positive Result		
							F26	No Health Consequences or Impact	A090804	False Positive Result		
					CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A090804	False Positive Result		
							F26	No Health Consequences or Impact	A090804	False Positive Result		
					CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A090804	False Positive Result		
							F26	No Health Consequences or Impact	A090804	False Positive Result		
					CT233	Surgical	E2403	No Clinical Signs, Symptoms or Conditions	A090804	False Positive Result		
							F26	No Health Consequences or Impact	A090804	False Positive Result		
					CT2406	Implantable	E2403	No Clinical Signs, Symptoms or Conditions	A090804	False Positive Result		
							F26	No Health Consequences or Impact	A090804	False Positive Result		
					CT2407	Active implantable	E2403	No Clinical Signs, Symptoms or Conditions	A090804	False Positive Result		
							F26	No Health Consequences or Impact	A090804	False Positive Result		
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A090804	False Positive Result		
							F26	No Health Consequences or Impact	A090804	False Positive Result		
					CT2727	Active implant devices/ancillaries	E2403	No Clinical Signs, Symptoms or Conditions	A090804	False Positive Result		
							F26	No Health Consequences or Impact	A090804	False Positive Result		
					CT2750	Implantable cardiac electrical stimulators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A090804	False Positive Result		
							F26	No Health Consequences or Impact	A090804	False Positive Result		
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A090804	False Positive Result		
							F26	No Health Consequences or Impact	A090804	False Positive Result		

					CT2985	Active	E2403	No Clinical Signs, Symptoms or Conditions	A090804	False Positive Result		
							F26	No Health Consequences or Impact	A090804	False Positive Result		
					CT3001	Implantable metal	E2403	No Clinical Signs, Symptoms or Conditions	A090804	False Positive Result		
							F26	No Health Consequences or Impact	A090804	False Positive Result		
					CT3002	Implantable synthetic polymer	E2403	No Clinical Signs, Symptoms or Conditions	A090804	False Positive Result		
							F26	No Health Consequences or Impact	A090804	False Positive Result		
					CT3006	Implantable inorganic material	E2403	No Clinical Signs, Symptoms or Conditions	A090804	False Positive Result		
							F26	No Health Consequences or Impact	A090804	False Positive Result		
					CT314	Electrically powered	E2403	No Clinical Signs, Symptoms or Conditions	A090804	False Positive Result		
							F26	No Health Consequences or Impact	A090804	False Positive Result		
					CT321	Long-term surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A090804	False Positive Result		
							F26	No Health Consequences or Impact	A090804	False Positive Result		
					CT449	Generators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A090804	False Positive Result		
							F26	No Health Consequences or Impact	A090804	False Positive Result		
					CT981	Single use	E2403	No Clinical Signs, Symptoms or Conditions	A090804	False Positive Result		
							F26	No Health Consequences or Impact	A090804	False Positive Result		
					CT983	Surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A090804	False Positive Result		
							F26	No Health Consequences or Impact	A090804	False Positive Result		
10231	2020	Death	47270	Cardiac resynchronization therapy implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A070912	Incorrect Interpretation of Signal	04/11/2020	04/11/2020
									A071301	Failure to Deliver Shock/Stimulation	04/11/2020	04/11/2020
									A25	No Apparent Adverse Event	04/11/2020	04/11/2020
							F21	Unexpected Deterioration	A070912	Incorrect Interpretation of Signal	04/11/2020	04/11/2020
									A071301	Failure to Deliver Shock/Stimulation	04/11/2020	04/11/2020
									A25	No Apparent Adverse Event	04/11/2020	04/11/2020
							F26	No Health Consequences or Impact	A070912	Incorrect Interpretation of Signal	04/11/2020	04/11/2020
									A071301	Failure to Deliver Shock/Stimulation	04/11/2020	04/11/2020
									A25	No Apparent Adverse Event	04/11/2020	04/11/2020
10232	2020	All other reportable incidents	36241	Pacemaker/defibrillator lead	CT1684	Cardiac pacemakers and associated devices	E1605	Cramp(s) /Muscle Spasm(s)	A051201	Device Dislodged or Dislocated	23/10/2020	23/10/2020
									A150206	Difficult to Insert	23/10/2020	23/10/2020
							F08	Hospitalization or Prolonged Hospitalization	A051201	Device Dislodged or Dislocated	23/10/2020	23/10/2020
									A150206	Difficult to Insert	23/10/2020	23/10/2020
							F12	Serious Injury/ Illness/ Impairment	A051201	Device Dislodged or Dislocated	23/10/2020	23/10/2020
									A150206	Difficult to Insert	23/10/2020	23/10/2020
							F1905	Device Revision or Replacement	A051201	Device Dislodged or Dislocated	23/10/2020	23/10/2020
									A150206	Difficult to Insert	23/10/2020	23/10/2020
10233	2020	All other reportable incidents	37265	Dual-chamber implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A150206	Difficult to Insert		
							F26	No Health Consequences or Impact	A150206	Difficult to Insert		
10234	2020	All other reportable incidents	47265	Dual-chamber implantable pacemaker, rate-responsive	CT141	Electrodes/Leads and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT145	Cardiovascular devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT233	Surgical	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT2406	Implantable	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT2407	Active implantable	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT2727	Active implant devices/ancillaries	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT2750	Implantable cardiac electrical stimulators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT2985	Active	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT3001	Implantable metal	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT3002	Implantable synthetic polymer	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT3006	Implantable inorganic material	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT314	Electrically powered	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT321	Long-term surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT449	Generators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT981	Single use	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT983	Surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
10235	2020	All other reportable incidents	37265	Dual-chamber implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	09/11/2020	09/11/2020
							F1901	Additional Surgery	A070504	Premature Discharge of Battery	09/11/2020	09/11/2020
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	09/11/2020	09/11/2020
10236	2020	All other reportable incidents	36241	Pacemaker/defibrillator lead	CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070909	Oversensing		
							F26	No Health Consequences or Impact	A070909	Oversensing		
10237	2020	All other reportable incidents	47265	Dual-chamber implantable pacemaker, rate-responsive	CT141	Electrodes/Leads and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070909	Oversensing		
							F26	No Health Consequences or Impact	A070909	Oversensing		
					CT145	Cardiovascular devices	E2403	No Clinical Signs, Symptoms or Conditions	A070909	Oversensing		
							F26	No Health Consequences or Impact	A070909	Oversensing		
					CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070909	Oversensing		
							F26	No Health Consequences or Impact	A070909	Oversensing		

					CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A070909	Oversensing		
							F26	No Health Consequences or Impact	A070909	Oversensing		
					CT233	Surgical	E2403	No Clinical Signs, Symptoms or Conditions	A070909	Oversensing		
							F26	No Health Consequences or Impact	A070909	Oversensing		
					CT2406	Implantable	E2403	No Clinical Signs, Symptoms or Conditions	A070909	Oversensing		
							F26	No Health Consequences or Impact	A070909	Oversensing		
					CT2407	Active implantable	E2403	No Clinical Signs, Symptoms or Conditions	A070909	Oversensing		
							F26	No Health Consequences or Impact	A070909	Oversensing		
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A070909	Oversensing		
							F26	No Health Consequences or Impact	A070909	Oversensing		
					CT2727	Active implant devices/ancillaries	E2403	No Clinical Signs, Symptoms or Conditions	A070909	Oversensing		
							F26	No Health Consequences or Impact	A070909	Oversensing		
					CT2750	Implantable cardiac electrical stimulators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070909	Oversensing		
							F26	No Health Consequences or Impact	A070909	Oversensing		
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A070909	Oversensing		
							F26	No Health Consequences or Impact	A070909	Oversensing		
					CT2985	Active	E2403	No Clinical Signs, Symptoms or Conditions	A070909	Oversensing		
							F26	No Health Consequences or Impact	A070909	Oversensing		
					CT3001	Implantable metal	E2403	No Clinical Signs, Symptoms or Conditions	A070909	Oversensing		
							F26	No Health Consequences or Impact	A070909	Oversensing		
					CT3002	Implantable synthetic polymer	E2403	No Clinical Signs, Symptoms or Conditions	A070909	Oversensing		
							F26	No Health Consequences or Impact	A070909	Oversensing		
					CT3006	Implantable inorganic material	E2403	No Clinical Signs, Symptoms or Conditions	A070909	Oversensing		
							F26	No Health Consequences or Impact	A070909	Oversensing		
					CT314	Electrically powered	E2403	No Clinical Signs, Symptoms or Conditions	A070909	Oversensing		
							F26	No Health Consequences or Impact	A070909	Oversensing		
					CT321	Long-term surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A070909	Oversensing		
							F26	No Health Consequences or Impact	A070909	Oversensing		
					CT449	Generators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070909	Oversensing		
							F26	No Health Consequences or Impact	A070909	Oversensing		
					CT981	Single use	E2403	No Clinical Signs, Symptoms or Conditions	A070909	Oversensing		
							F26	No Health Consequences or Impact	A070909	Oversensing		
					CT983	Surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A070909	Oversensing		
							F26	No Health Consequences or Impact	A070909	Oversensing		
10238	2020	All other reportable incidents	35852	Single-chamber implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E2104	Electric Shock	A070909	Oversensing		
									A071302	Inappropriate/Inadequate Shock/Stimulation		
							F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A070909	Oversensing		
									A071302	Inappropriate/Inadequate Shock/Stimulation		
							F24	Insufficient Health Impact Information	A070909	Oversensing		
									A071302	Inappropriate/Inadequate Shock/Stimulation		
10239	2020	All other reportable incidents	36241	Pacemaker/defibrillator lead	CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070102	High Capture Threshold	04/11/2020	04/11/2020
									A070909	Oversensing	04/11/2020	04/11/2020
									A072202	Low Impedance	04/11/2020	04/11/2020
									A090801	Signal Artifact/Noise	04/11/2020	04/11/2020
							F19	Surgical Intervention	A070102	High Capture Threshold	04/11/2020	04/11/2020
									A070909	Oversensing	04/11/2020	04/11/2020
									A072202	Low Impedance	04/11/2020	04/11/2020
									A090801	Signal Artifact/Noise	04/11/2020	04/11/2020
10240	2020	All other reportable incidents	47265	Dual-chamber implantable pacemaker, rate-responsive	CT141	Electrodes/Leads and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery		
							F24	Insufficient Health Impact Information	A070504	Premature Discharge of Battery		
					CT145	Cardiovascular devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery		
							F24	Insufficient Health Impact Information	A070504	Premature Discharge of Battery		
					CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery		
							F24	Insufficient Health Impact Information	A070504	Premature Discharge of Battery		
					CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery		
							F24	Insufficient Health Impact Information	A070504	Premature Discharge of Battery		
					CT233	Surgical	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery		
							F24	Insufficient Health Impact Information	A070504	Premature Discharge of Battery		
					CT2406	Implantable	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery		
							F24	Insufficient Health Impact Information	A070504	Premature Discharge of Battery		
					CT2407	Active implantable	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery		
							F24	Insufficient Health Impact Information	A070504	Premature Discharge of Battery		
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery		
							F24	Insufficient Health Impact Information	A070504	Premature Discharge of Battery		
					CT2727	Active implant devices/ancillaries	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery		
							F24	Insufficient Health Impact Information	A070504	Premature Discharge of Battery		
					CT2750	Implantable cardiac electrical stimulators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery		
							F24	Insufficient Health Impact Information	A070504	Premature Discharge of Battery		
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery		
							F24	Insufficient Health Impact Information	A070504	Premature Discharge of Battery		
					CT2985	Active	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery		
							F24	Insufficient Health Impact Information	A070504	Premature Discharge of Battery		
					CT3001	Implantable metal	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery		
							F24	Insufficient Health Impact Information	A070504	Premature Discharge of Battery		
					CT3002	Implantable synthetic polymer	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		

						F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery		
						F24	Insufficient Health Impact Information	A070504	Premature Discharge of Battery		
					CT3006	Implantable inorganic material	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery		
							F24	Insufficient Health Impact Information	A070504	Premature Discharge of Battery		
					CT314	Electrically powered	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery		
							F24	Insufficient Health Impact Information	A070504	Premature Discharge of Battery		
					CT321	Long-term surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery		
							F24	Insufficient Health Impact Information	A070504	Premature Discharge of Battery		
					CT449	Generators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery		
							F24	Insufficient Health Impact Information	A070504	Premature Discharge of Battery		
					CT981	Single use	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery		
							F24	Insufficient Health Impact Information	A070504	Premature Discharge of Battery		
					CT983	Surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery		
							F24	Insufficient Health Impact Information	A070504	Premature Discharge of Battery		
10241	2020	All other reportable incidents	47267	Single-chamber implantable pacemaker, rate-responsive	CT141	Electrodes/Leads and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	23/10/2020	23/10/2020
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	23/10/2020	23/10/2020
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	23/10/2020	23/10/2020
					CT145	Cardiovascular devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	23/10/2020	23/10/2020
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	23/10/2020	23/10/2020
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	23/10/2020	23/10/2020
					CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	23/10/2020	23/10/2020
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	23/10/2020	23/10/2020
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	23/10/2020	23/10/2020
					CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	23/10/2020	23/10/2020
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	23/10/2020	23/10/2020
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	23/10/2020	23/10/2020
					CT233	Surgical	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	23/10/2020	23/10/2020
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	23/10/2020	23/10/2020
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	23/10/2020	23/10/2020
					CT2406	Implantable	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	23/10/2020	23/10/2020
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	23/10/2020	23/10/2020
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	23/10/2020	23/10/2020
					CT2407	Active implantable	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	23/10/2020	23/10/2020
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	23/10/2020	23/10/2020
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	23/10/2020	23/10/2020
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	23/10/2020	23/10/2020
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	23/10/2020	23/10/2020
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	23/10/2020	23/10/2020
					CT2727	Active implant devices/ancillaries	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	23/10/2020	23/10/2020
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	23/10/2020	23/10/2020
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	23/10/2020	23/10/2020
					CT2750	Implantable cardiac electrical stimulators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	23/10/2020	23/10/2020
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	23/10/2020	23/10/2020
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	23/10/2020	23/10/2020
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	23/10/2020	23/10/2020
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	23/10/2020	23/10/2020
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	23/10/2020	23/10/2020
					CT2985	Active	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	23/10/2020	23/10/2020
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	23/10/2020	23/10/2020
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	23/10/2020	23/10/2020
					CT3001	Implantable metal	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	23/10/2020	23/10/2020
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	23/10/2020	23/10/2020
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	23/10/2020	23/10/2020
					CT3002	Implantable synthetic polymer	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	23/10/2020	23/10/2020
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	23/10/2020	23/10/2020
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	23/10/2020	23/10/2020
					CT3006	Implantable inorganic material	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	23/10/2020	23/10/2020
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	23/10/2020	23/10/2020
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	23/10/2020	23/10/2020
					CT314	Electrically powered	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	23/10/2020	23/10/2020
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	23/10/2020	23/10/2020
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	23/10/2020	23/10/2020
					CT321	Long-term surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	23/10/2020	23/10/2020
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	23/10/2020	23/10/2020
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	23/10/2020	23/10/2020
					CT449	Generators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	23/10/2020	23/10/2020
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	23/10/2020	23/10/2020
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	23/10/2020	23/10/2020
					CT981	Single use	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	23/10/2020	23/10/2020
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	23/10/2020	23/10/2020
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	23/10/2020	23/10/2020
					CT983	Surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	23/10/2020	23/10/2020
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	23/10/2020	23/10/2020
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	23/10/2020	23/10/2020
10242	2020	All other reportable incidents	35852	Single-chamber implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A070909	Oversensing	14/10/2020	14/10/2020
							F1905	Device Revision or Replacement	A070909	Oversensing	14/10/2020	14/10/2020
10243	2020	All other reportable incidents	35852	Single-chamber implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E2104	Electric Shock	A071302	Inappropriate/Inadequate Shock/Stimulation		
							F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A071302	Inappropriate/Inadequate Shock/Stimulation		
							F26	No Health Consequences or Impact	A071302	Inappropriate/Inadequate Shock/Stimulation		
10244	2020	All other reportable incidents	47263	Cardiac resynchronization therapy implantable pacemaker	CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A0908	Incorrect, Inadequate or Imprecise Result or Readings	14/10/2020	14/10/2020

							F08	Hospitalization or Prolonged Hospitalization	A0908	Incorrect, Inadequate or Imprecise Result or Readings	14/10/2020	14/10/2020
							F1905	Device Revision or Replacement	A0908	Incorrect, Inadequate or Imprecise Result or Readings	14/10/2020	14/10/2020
					CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A0908	Incorrect, Inadequate or Imprecise Result or Readings	14/10/2020	14/10/2020
							F08	Hospitalization or Prolonged Hospitalization	A0908	Incorrect, Inadequate or Imprecise Result or Readings	14/10/2020	14/10/2020
							F1905	Device Revision or Replacement	A0908	Incorrect, Inadequate or Imprecise Result or Readings	14/10/2020	14/10/2020
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A0908	Incorrect, Inadequate or Imprecise Result or Readings	14/10/2020	14/10/2020
							F08	Hospitalization or Prolonged Hospitalization	A0908	Incorrect, Inadequate or Imprecise Result or Readings	14/10/2020	14/10/2020
							F1905	Device Revision or Replacement	A0908	Incorrect, Inadequate or Imprecise Result or Readings	14/10/2020	14/10/2020
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A0908	Incorrect, Inadequate or Imprecise Result or Readings	14/10/2020	14/10/2020
							F08	Hospitalization or Prolonged Hospitalization	A0908	Incorrect, Inadequate or Imprecise Result or Readings	14/10/2020	14/10/2020
							F1905	Device Revision or Replacement	A0908	Incorrect, Inadequate or Imprecise Result or Readings	14/10/2020	14/10/2020
10245	2020	All other reportable incidents	37265	Dual-chamber implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E2104	Electric Shock	A071302	Inappropriate/Inadequate Shock/Stimulation		
							F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A071302	Inappropriate/Inadequate Shock/Stimulation		
							F1905	Device Revision or Replacement	A071302	Inappropriate/Inadequate Shock/Stimulation		
							F26	No Health Consequences or Impact	A071302	Inappropriate/Inadequate Shock/Stimulation		
10246	2020	All other reportable incidents	17882	Physiologic-monitoring defibrillation system	CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070802	Intermittent Loss of Power		
							F27	Problem Identified During Non-Clinical Procedure	A070802	Intermittent Loss of Power		
10247	2020	All other reportable incidents	47265	Dual-chamber implantable pacemaker, rate-responsive	CT141	Electrodes/Leads and associated devices	E2401	Insufficient Information	A070504	Premature Discharge of Battery	25/01/2021	25/01/2021
							E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	25/01/2021	25/01/2021
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	25/01/2021	25/01/2021
							F1903	Device Explantation	A070504	Premature Discharge of Battery	25/01/2021	25/01/2021
					CT145	Cardiovascular devices	E2401	Insufficient Information	A070504	Premature Discharge of Battery	25/01/2021	25/01/2021
							E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	25/01/2021	25/01/2021
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	25/01/2021	25/01/2021
							F1903	Device Explantation	A070504	Premature Discharge of Battery	25/01/2021	25/01/2021
					CT1684	Cardiac pacemakers and associated devices	E2401	Insufficient Information	A070504	Premature Discharge of Battery	25/01/2021	25/01/2021
							E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	25/01/2021	25/01/2021
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	25/01/2021	25/01/2021
							F1903	Device Explantation	A070504	Premature Discharge of Battery	25/01/2021	25/01/2021
					CT1685	Cardiac pacemakers	E2401	Insufficient Information	A070504	Premature Discharge of Battery	25/01/2021	25/01/2021
							E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	25/01/2021	25/01/2021
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	25/01/2021	25/01/2021
							F1903	Device Explantation	A070504	Premature Discharge of Battery	25/01/2021	25/01/2021
					CT233	Surgical	E2401	Insufficient Information	A070504	Premature Discharge of Battery	25/01/2021	25/01/2021
							E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	25/01/2021	25/01/2021
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	25/01/2021	25/01/2021
							F1903	Device Explantation	A070504	Premature Discharge of Battery	25/01/2021	25/01/2021
					CT2406	Implantable	E2401	Insufficient Information	A070504	Premature Discharge of Battery	25/01/2021	25/01/2021
							E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	25/01/2021	25/01/2021
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	25/01/2021	25/01/2021
							F1903	Device Explantation	A070504	Premature Discharge of Battery	25/01/2021	25/01/2021
					CT2407	Active implantable	E2401	Insufficient Information	A070504	Premature Discharge of Battery	25/01/2021	25/01/2021
							E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	25/01/2021	25/01/2021
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	25/01/2021	25/01/2021
							F1903	Device Explantation	A070504	Premature Discharge of Battery	25/01/2021	25/01/2021
					CT257	Implantable cardiac pacemakers	E2401	Insufficient Information	A070504	Premature Discharge of Battery	25/01/2021	25/01/2021
							E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	25/01/2021	25/01/2021
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	25/01/2021	25/01/2021
							F1903	Device Explantation	A070504	Premature Discharge of Battery	25/01/2021	25/01/2021
					CT2727	Active implant devices/ancillaries	E2401	Insufficient Information	A070504	Premature Discharge of Battery	25/01/2021	25/01/2021
							E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	25/01/2021	25/01/2021
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	25/01/2021	25/01/2021
							F1903	Device Explantation	A070504	Premature Discharge of Battery	25/01/2021	25/01/2021
					CT2750	Implantable cardiac electrical stimulators and associated devices	E2401	Insufficient Information	A070504	Premature Discharge of Battery	25/01/2021	25/01/2021
							E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	25/01/2021	25/01/2021
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	25/01/2021	25/01/2021
							F1903	Device Explantation	A070504	Premature Discharge of Battery	25/01/2021	25/01/2021
					CT2866	Implantable cardiac electrical stimulators	E2401	Insufficient Information	A070504	Premature Discharge of Battery	25/01/2021	25/01/2021
							E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	25/01/2021	25/01/2021
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	25/01/2021	25/01/2021
							F1903	Device Explantation	A070504	Premature Discharge of Battery	25/01/2021	25/01/2021
					CT2985	Active	E2401	Insufficient Information	A070504	Premature Discharge of Battery	25/01/2021	25/01/2021
							E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	25/01/2021	25/01/2021
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	25/01/2021	25/01/2021
							F1903	Device Explantation	A070504	Premature Discharge of Battery	25/01/2021	25/01/2021
					CT3001	Implantable metal	E2401	Insufficient Information	A070504	Premature Discharge of Battery	25/01/2021	25/01/2021
							E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	25/01/2021	25/01/2021
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	25/01/2021	25/01/2021
							F1903	Device Explantation	A070504	Premature Discharge of Battery	25/01/2021	25/01/2021
					CT3002	Implantable synthetic polymer	E2401	Insufficient Information	A070504	Premature Discharge of Battery	25/01/2021	25/01/2021
							E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	25/01/2021	25/01/2021
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	25/01/2021	25/01/2021
							F1903	Device Explantation	A070504	Premature Discharge of Battery	25/01/2021	25/01/2021
					CT3006	Implantable inorganic material	E2401	Insufficient Information	A070504	Premature Discharge of Battery	25/01/2021	25/01/2021
							E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	25/01/2021	25/01/2021
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	25/01/2021	25/01/2021
							F1903	Device Explantation	A070504	Premature Discharge of Battery	25/01/2021	25/01/2021
					CT314	Electrically powered	E2401	Insufficient Information	A070504	Premature Discharge of Battery	25/01/2021	25/01/2021
							E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	25/01/2021	25/01/2021

							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	25/01/2021	25/01/2021
							F1903	Device Explantation	A070504	Premature Discharge of Battery	25/01/2021	25/01/2021
					CT321	Long-term surgical invasive	E2401	Insufficient Information	A070504	Premature Discharge of Battery	25/01/2021	25/01/2021
							E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	25/01/2021	25/01/2021
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	25/01/2021	25/01/2021
							F1903	Device Explantation	A070504	Premature Discharge of Battery	25/01/2021	25/01/2021
					CT449	Generators and associated devices	E2401	Insufficient Information	A070504	Premature Discharge of Battery	25/01/2021	25/01/2021
							E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	25/01/2021	25/01/2021
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	25/01/2021	25/01/2021
							F1903	Device Explantation	A070504	Premature Discharge of Battery	25/01/2021	25/01/2021
					CT981	Single use	E2401	Insufficient Information	A070504	Premature Discharge of Battery	25/01/2021	25/01/2021
							E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	25/01/2021	25/01/2021
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	25/01/2021	25/01/2021
							F1903	Device Explantation	A070504	Premature Discharge of Battery	25/01/2021	25/01/2021
					CT983	Surgical invasive	E2401	Insufficient Information	A070504	Premature Discharge of Battery	25/01/2021	25/01/2021
							E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	25/01/2021	25/01/2021
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	25/01/2021	25/01/2021
							F1903	Device Explantation	A070504	Premature Discharge of Battery	25/01/2021	25/01/2021
10248	2020	All other reportable incidents	35852	Single-chamber implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	25/06/2021	25/06/2021
									A070909	Oversensing	25/06/2021	25/06/2021
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	25/06/2021	25/06/2021
									A070909	Oversensing	25/06/2021	25/06/2021
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	25/06/2021	25/06/2021
									A070909	Oversensing	25/06/2021	25/06/2021
10249	2020	All other reportable incidents	47265	Dual-chamber implantable pacemaker, rate-responsive	CT141	Electrodes/Leads and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A0712	Pacing Problem		
							F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A0712	Pacing Problem		
							F26	No Health Consequences or Impact	A0712	Pacing Problem		
					CT145	Cardiovascular devices	E2403	No Clinical Signs, Symptoms or Conditions	A0712	Pacing Problem		
							F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A0712	Pacing Problem		
							F26	No Health Consequences or Impact	A0712	Pacing Problem		
					CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A0712	Pacing Problem		
							F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A0712	Pacing Problem		
							F26	No Health Consequences or Impact	A0712	Pacing Problem		
					CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A0712	Pacing Problem		
							F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A0712	Pacing Problem		
							F26	No Health Consequences or Impact	A0712	Pacing Problem		
					CT233	Surgical	E2403	No Clinical Signs, Symptoms or Conditions	A0712	Pacing Problem		
							F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A0712	Pacing Problem		
							F26	No Health Consequences or Impact	A0712	Pacing Problem		
					CT2406	Implantable	E2403	No Clinical Signs, Symptoms or Conditions	A0712	Pacing Problem		
							F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A0712	Pacing Problem		
							F26	No Health Consequences or Impact	A0712	Pacing Problem		
					CT2407	Active implantable	E2403	No Clinical Signs, Symptoms or Conditions	A0712	Pacing Problem		
							F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A0712	Pacing Problem		
							F26	No Health Consequences or Impact	A0712	Pacing Problem		
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A0712	Pacing Problem		
							F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A0712	Pacing Problem		
							F26	No Health Consequences or Impact	A0712	Pacing Problem		
					CT2727	Active implant devices/ancillaries	E2403	No Clinical Signs, Symptoms or Conditions	A0712	Pacing Problem		
							F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A0712	Pacing Problem		
							F26	No Health Consequences or Impact	A0712	Pacing Problem		
					CT2750	Implantable cardiac electrical stimulators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A0712	Pacing Problem		
							F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A0712	Pacing Problem		
							F26	No Health Consequences or Impact	A0712	Pacing Problem		
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A0712	Pacing Problem		
							F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A0712	Pacing Problem		
							F26	No Health Consequences or Impact	A0712	Pacing Problem		
					CT2985	Active	E2403	No Clinical Signs, Symptoms or Conditions	A0712	Pacing Problem		
							F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A0712	Pacing Problem		
							F26	No Health Consequences or Impact	A0712	Pacing Problem		
					CT3001	Implantable metal	E2403	No Clinical Signs, Symptoms or Conditions	A0712	Pacing Problem		
							F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A0712	Pacing Problem		
							F26	No Health Consequences or Impact	A0712	Pacing Problem		
					CT3002	Implantable synthetic polymer	E2403	No Clinical Signs, Symptoms or Conditions	A0712	Pacing Problem		
							F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A0712	Pacing Problem		
							F26	No Health Consequences or Impact	A0712	Pacing Problem		
					CT3006	Implantable inorganic material	E2403	No Clinical Signs, Symptoms or Conditions	A0712	Pacing Problem		
							F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A0712	Pacing Problem		
							F26	No Health Consequences or Impact	A0712	Pacing Problem		
					CT314	Electrically powered	E2403	No Clinical Signs, Symptoms or Conditions	A0712	Pacing Problem		
							F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A0712	Pacing Problem		
							F26	No Health Consequences or Impact	A0712	Pacing Problem		
					CT321	Long-term surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A0712	Pacing Problem		
							F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A0712	Pacing Problem		
							F26	No Health Consequences or Impact	A0712	Pacing Problem		
					CT449	Generators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A0712	Pacing Problem		
							F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A0712	Pacing Problem		
							F26	No Health Consequences or Impact	A0712	Pacing Problem		
					CT981	Single use	E2403	No Clinical Signs, Symptoms or Conditions	A0712	Pacing Problem		
							F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A0712	Pacing Problem		
							F26	No Health Consequences or Impact	A0712	Pacing Problem		
					CT983	Surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A0712	Pacing Problem		
							F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A0712	Pacing Problem		
							F26	No Health Consequences or Impact	A0712	Pacing Problem		
10250	2020	All other reportable incidents	47270	Cardiac resynchronization therapy implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E2401	Insufficient Information	A160601	Premature Elective Replacement Indicator		
							E2403	No Clinical Signs, Symptoms or Conditions	A160601	Premature Elective Replacement Indicator		
							F24	Insufficient Health Impact Information	A160601	Premature Elective Replacement Indicator		
10251	2020	All other reportable incidents	47267	Single-chamber implantable pacemaker, rate-	CT141	Electrodes/Leads and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		

				responsive			F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT145	Cardiovascular devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT233	Surgical	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT2406	Implantable	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT2407	Active implantable	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT2727	Active implant devices/ancillaries	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT2750	Implantable cardiac electrical stimulators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT2985	Active	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT3001	Implantable metal	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT3002	Implantable synthetic polymer	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT3006	Implantable inorganic material	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT314	Electrically powered	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT321	Long-term surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT449	Generators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT981	Single use	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT983	Surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
10252	2020	All other reportable incidents	17882	Physiologic-monitoring defibrillation system	CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A07	Electrical /Electronic Property Problem		
							F26	No Health Consequences or Impact	A07	Electrical /Electronic Property Problem		
10253	2020	All other reportable incidents	47270	Cardiac resynchronization therapy implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A070909	Oversensing		
							F26	No Health Consequences or Impact	A070909	Oversensing		
10254	2020	All other reportable incidents	47267	Single-chamber implantable pacemaker, rate-responsive	CT141	Electrodes/Leads and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F02	Death	A070504	Premature Discharge of Battery		
					CT145	Cardiovascular devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F02	Death	A070504	Premature Discharge of Battery		
					CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F02	Death	A070504	Premature Discharge of Battery		
					CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F02	Death	A070504	Premature Discharge of Battery		
					CT233	Surgical	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F02	Death	A070504	Premature Discharge of Battery		
					CT2406	Implantable	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F02	Death	A070504	Premature Discharge of Battery		
					CT2407	Active implantable	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F02	Death	A070504	Premature Discharge of Battery		
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F02	Death	A070504	Premature Discharge of Battery		
					CT2727	Active implant devices/ancillaries	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F02	Death	A070504	Premature Discharge of Battery		
					CT2750	Implantable cardiac electrical stimulators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F02	Death	A070504	Premature Discharge of Battery		
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F02	Death	A070504	Premature Discharge of Battery		
					CT2985	Active	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F02	Death	A070504	Premature Discharge of Battery		
					CT3001	Implantable metal	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		

							F02	Death	A070504	Premature Discharge of Battery		
					CT3002	Implantable synthetic polymer	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F02	Death	A070504	Premature Discharge of Battery		
					CT3006	Implantable inorganic material	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F02	Death	A070504	Premature Discharge of Battery		
					CT314	Electrically powered	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F02	Death	A070504	Premature Discharge of Battery		
					CT321	Long-term surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F02	Death	A070504	Premature Discharge of Battery		
					CT449	Generators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F02	Death	A070504	Premature Discharge of Battery		
					CT981	Single use	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F02	Death	A070504	Premature Discharge of Battery		
					CT983	Surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F02	Death	A070504	Premature Discharge of Battery		
10255	2020	All other reportable incidents	47265	Dual-chamber implantable pacemaker, rate-responsive	CT141	Electrodes/Leads and associated devices	E011903	Syncope/Fainting	A070909	Oversensing		
									A071204	No Pacing		
							E0623	Unspecified Heart Problem	A070909	Oversensing		
									A071204	No Pacing		
							E1603	Bone Fracture(s)	A070909	Oversensing		
									A071204	No Pacing		
							E2402	Appropriate Clinical Signs, Symptoms and Conditions Term/Code not Available	A070909	Oversensing		
									A071204	No Pacing		
							F23	Unexpected Medical Intervention	A070909	Oversensing		
									A071204	No Pacing		
					CT145	Cardiovascular devices	E011903	Syncope/Fainting	A070909	Oversensing		
									A071204	No Pacing		
							E0623	Unspecified Heart Problem	A070909	Oversensing		
									A071204	No Pacing		
							E1603	Bone Fracture(s)	A070909	Oversensing		
									A071204	No Pacing		
							E2402	Appropriate Clinical Signs, Symptoms and Conditions Term/Code not Available	A070909	Oversensing		
									A071204	No Pacing		
							F23	Unexpected Medical Intervention	A070909	Oversensing		
									A071204	No Pacing		
					CT1684	Cardiac pacemakers and associated devices	E011903	Syncope/Fainting	A070909	Oversensing		
									A071204	No Pacing		
							E0623	Unspecified Heart Problem	A070909	Oversensing		
									A071204	No Pacing		
							E1603	Bone Fracture(s)	A070909	Oversensing		
									A071204	No Pacing		
							E2402	Appropriate Clinical Signs, Symptoms and Conditions Term/Code not Available	A070909	Oversensing		
									A071204	No Pacing		
							F23	Unexpected Medical Intervention	A070909	Oversensing		
									A071204	No Pacing		
					CT1685	Cardiac pacemakers	E011903	Syncope/Fainting	A070909	Oversensing		
									A071204	No Pacing		
							E0623	Unspecified Heart Problem	A070909	Oversensing		
									A071204	No Pacing		
							E1603	Bone Fracture(s)	A070909	Oversensing		
									A071204	No Pacing		
							E2402	Appropriate Clinical Signs, Symptoms and Conditions Term/Code not Available	A070909	Oversensing		
									A071204	No Pacing		
							F23	Unexpected Medical Intervention	A070909	Oversensing		
									A071204	No Pacing		
					CT233	Surgical	E011903	Syncope/Fainting	A070909	Oversensing		
									A071204	No Pacing		
							E0623	Unspecified Heart Problem	A070909	Oversensing		
									A071204	No Pacing		
							E1603	Bone Fracture(s)	A070909	Oversensing		
									A071204	No Pacing		
							E2402	Appropriate Clinical Signs, Symptoms and Conditions Term/Code not Available	A070909	Oversensing		
									A071204	No Pacing		
							F23	Unexpected Medical Intervention	A070909	Oversensing		
									A071204	No Pacing		
					CT2406	Implantable	E011903	Syncope/Fainting	A070909	Oversensing		
									A071204	No Pacing		
							E0623	Unspecified Heart Problem	A070909	Oversensing		
									A071204	No Pacing		
							E1603	Bone Fracture(s)	A070909	Oversensing		
									A071204	No Pacing		
							E2402	Appropriate Clinical Signs, Symptoms and Conditions Term/Code not Available	A070909	Oversensing		
									A071204	No Pacing		
							F23	Unexpected Medical Intervention	A070909	Oversensing		
									A071204	No Pacing		
					CT2407	Active implantable	E011903	Syncope/Fainting	A070909	Oversensing		
									A071204	No Pacing		
							E0623	Unspecified Heart Problem	A070909	Oversensing		
									A071204	No Pacing		
							E1603	Bone Fracture(s)	A070909	Oversensing		
									A071204	No Pacing		
							E2402	Appropriate Clinical Signs, Symptoms and Conditions Term/Code not Available	A070909	Oversensing		
									A071204	No Pacing		
							F23	Unexpected Medical Intervention	A070909	Oversensing		
									A071204	No Pacing		
					CT257	Implantable cardiac pacemakers	E011903	Syncope/Fainting	A070909	Oversensing		
									A071204	No Pacing		
							E0623	Unspecified Heart Problem	A070909	Oversensing		

									A071204	No Pacing		
							E1603	Bone Fracture(s)	A070909	Oversensing		
									A071204	No Pacing		
							E2402	Appropriate Clinical Signs, Symptoms and Conditions Term/Code not Available	A070909	Oversensing		
									A071204	No Pacing		
							F23	Unexpected Medical Intervention	A070909	Oversensing		
									A071204	No Pacing		
					CT2727	Active implant devices/ancillaries	E011903	Syncope/Fainting	A070909	Oversensing		
									A071204	No Pacing		
							E0623	Unspecified Heart Problem	A070909	Oversensing		
									A071204	No Pacing		
							E1603	Bone Fracture(s)	A070909	Oversensing		
									A071204	No Pacing		
							E2402	Appropriate Clinical Signs, Symptoms and Conditions Term/Code not Available	A070909	Oversensing		
									A071204	No Pacing		
							F23	Unexpected Medical Intervention	A070909	Oversensing		
									A071204	No Pacing		
					CT2750	Implantable cardiac electrical stimulators and associated devices	E011903	Syncope/Fainting	A070909	Oversensing		
									A071204	No Pacing		
							E0623	Unspecified Heart Problem	A070909	Oversensing		
									A071204	No Pacing		
							E1603	Bone Fracture(s)	A070909	Oversensing		
									A071204	No Pacing		
							E2402	Appropriate Clinical Signs, Symptoms and Conditions Term/Code not Available	A070909	Oversensing		
									A071204	No Pacing		
							F23	Unexpected Medical Intervention	A070909	Oversensing		
									A071204	No Pacing		
					CT2866	Implantable cardiac electrical stimulators	E011903	Syncope/Fainting	A070909	Oversensing		
									A071204	No Pacing		
							E0623	Unspecified Heart Problem	A070909	Oversensing		
									A071204	No Pacing		
							E1603	Bone Fracture(s)	A070909	Oversensing		
									A071204	No Pacing		
							E2402	Appropriate Clinical Signs, Symptoms and Conditions Term/Code not Available	A070909	Oversensing		
									A071204	No Pacing		
							F23	Unexpected Medical Intervention	A070909	Oversensing		
									A071204	No Pacing		
					CT2985	Active	E011903	Syncope/Fainting	A070909	Oversensing		
									A071204	No Pacing		
							E0623	Unspecified Heart Problem	A070909	Oversensing		
									A071204	No Pacing		
							E1603	Bone Fracture(s)	A070909	Oversensing		
									A071204	No Pacing		
							E2402	Appropriate Clinical Signs, Symptoms and Conditions Term/Code not Available	A070909	Oversensing		
									A071204	No Pacing		
							F23	Unexpected Medical Intervention	A070909	Oversensing		
									A071204	No Pacing		
					CT3001	Implantable metal	E011903	Syncope/Fainting	A070909	Oversensing		
									A071204	No Pacing		
							E0623	Unspecified Heart Problem	A070909	Oversensing		
									A071204	No Pacing		
							E1603	Bone Fracture(s)	A070909	Oversensing		
									A071204	No Pacing		
							E2402	Appropriate Clinical Signs, Symptoms and Conditions Term/Code not Available	A070909	Oversensing		
									A071204	No Pacing		
							F23	Unexpected Medical Intervention	A070909	Oversensing		
									A071204	No Pacing		
					CT3002	Implantable synthetic polymer	E011903	Syncope/Fainting	A070909	Oversensing		
									A071204	No Pacing		
							E0623	Unspecified Heart Problem	A070909	Oversensing		
									A071204	No Pacing		
							E1603	Bone Fracture(s)	A070909	Oversensing		
									A071204	No Pacing		
							E2402	Appropriate Clinical Signs, Symptoms and Conditions Term/Code not Available	A070909	Oversensing		
									A071204	No Pacing		
							F23	Unexpected Medical Intervention	A070909	Oversensing		
									A071204	No Pacing		
					CT3006	Implantable inorganic material	E011903	Syncope/Fainting	A070909	Oversensing		
									A071204	No Pacing		
							E0623	Unspecified Heart Problem	A070909	Oversensing		
									A071204	No Pacing		
							E1603	Bone Fracture(s)	A070909	Oversensing		
									A071204	No Pacing		
							E2402	Appropriate Clinical Signs, Symptoms and Conditions Term/Code not Available	A070909	Oversensing		
									A071204	No Pacing		
							F23	Unexpected Medical Intervention	A070909	Oversensing		
									A071204	No Pacing		
					CT314	Electrically powered	E011903	Syncope/Fainting	A070909	Oversensing		
									A071204	No Pacing		
							E0623	Unspecified Heart Problem	A070909	Oversensing		
									A071204	No Pacing		
							E1603	Bone Fracture(s)	A070909	Oversensing		
									A071204	No Pacing		
							E2402	Appropriate Clinical Signs, Symptoms and Conditions Term/Code not Available	A070909	Oversensing		
									A071204	No Pacing		
							F23	Unexpected Medical Intervention	A070909	Oversensing		
									A071204	No Pacing		
					CT321	Long-term surgical invasive	E011903	Syncope/Fainting	A070909	Oversensing		

									A071204	No Pacing		
							E0623	Unspecified Heart Problem	A070909	Oversensing		
									A071204	No Pacing		
							E1603	Bone Fracture(s)	A070909	Oversensing		
									A071204	No Pacing		
							E2402	Appropriate Clinical Signs, Symptoms and Conditions Term/Code not Available	A070909	Oversensing		
									A071204	No Pacing		
							F23	Unexpected Medical Intervention	A070909	Oversensing		
									A071204	No Pacing		
					CT449	Generators and associated devices	E011903	Syncope/Fainting	A070909	Oversensing		
									A071204	No Pacing		
							E0623	Unspecified Heart Problem	A070909	Oversensing		
									A071204	No Pacing		
							E1603	Bone Fracture(s)	A070909	Oversensing		
									A071204	No Pacing		
							E2402	Appropriate Clinical Signs, Symptoms and Conditions Term/Code not Available	A070909	Oversensing		
									A071204	No Pacing		
							F23	Unexpected Medical Intervention	A070909	Oversensing		
									A071204	No Pacing		
					CT981	Single use	E011903	Syncope/Fainting	A070909	Oversensing		
									A071204	No Pacing		
							E0623	Unspecified Heart Problem	A070909	Oversensing		
									A071204	No Pacing		
							E1603	Bone Fracture(s)	A070909	Oversensing		
									A071204	No Pacing		
							E2402	Appropriate Clinical Signs, Symptoms and Conditions Term/Code not Available	A070909	Oversensing		
									A071204	No Pacing		
							F23	Unexpected Medical Intervention	A070909	Oversensing		
									A071204	No Pacing		
					CT983	Surgical invasive	E011903	Syncope/Fainting	A070909	Oversensing		
									A071204	No Pacing		
							E0623	Unspecified Heart Problem	A070909	Oversensing		
									A071204	No Pacing		
							E1603	Bone Fracture(s)	A070909	Oversensing		
									A071204	No Pacing		
							E2402	Appropriate Clinical Signs, Symptoms and Conditions Term/Code not Available	A070909	Oversensing		
									A071204	No Pacing		
							F23	Unexpected Medical Intervention	A070909	Oversensing		
									A071204	No Pacing		
10256	2020	All other reportable incidents	35223	Endocardial/interventricular septal pacing lead	CT1684	Cardiac pacemakers and associated devices	E0112	Dizziness	A070103	Intermittent Capture		
							F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A070103	Intermittent Capture		
							F11	Non-Serious Injury/ Illness/ Impairment	A070103	Intermittent Capture		
							F1905	Device Revision or Replacement	A070103	Intermittent Capture		
10257	2020	All other reportable incidents	35852	Single-chamber implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	12/11/2020	12/11/2020
							F19	Surgical Intervention	A070504	Premature Discharge of Battery	12/11/2020	12/11/2020
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	12/11/2020	12/11/2020
10258	2020	All other reportable incidents	47263	Cardiac resynchronization therapy implantable pacemaker	CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070909	Oversensing		
							F26	No Health Consequences or Impact	A070909	Oversensing		
					CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A070909	Oversensing		
							F26	No Health Consequences or Impact	A070909	Oversensing		
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A070909	Oversensing		
							F26	No Health Consequences or Impact	A070909	Oversensing		
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A070909	Oversensing		
							F26	No Health Consequences or Impact	A070909	Oversensing		
10259	2020	All other reportable incidents	37265	Dual-chamber implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E060102	Atrial Fibrillation	A0712	Pacing Problem		
									A0909	Unexpected Therapeutic Results		
							E2403	No Clinical Signs, Symptoms or Conditions	A0712	Pacing Problem		
									A0909	Unexpected Therapeutic Results		
							F26	No Health Consequences or Impact	A0712	Pacing Problem		
									A0909	Unexpected Therapeutic Results		
10260	2020	All other reportable incidents	47265	Dual-chamber implantable pacemaker, rate-responsive	CT141	Electrodes/Leads and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	10/12/2020	10/12/2020
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	10/12/2020	10/12/2020
							F12	Serious Injury/ Illness/ Impairment	A070504	Premature Discharge of Battery	10/12/2020	10/12/2020
							F1903	Device Explantation	A070504	Premature Discharge of Battery	10/12/2020	10/12/2020
					CT145	Cardiovascular devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	10/12/2020	10/12/2020
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	10/12/2020	10/12/2020
							F12	Serious Injury/ Illness/ Impairment	A070504	Premature Discharge of Battery	10/12/2020	10/12/2020
							F1903	Device Explantation	A070504	Premature Discharge of Battery	10/12/2020	10/12/2020
					CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	10/12/2020	10/12/2020
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	10/12/2020	10/12/2020
							F12	Serious Injury/ Illness/ Impairment	A070504	Premature Discharge of Battery	10/12/2020	10/12/2020
							F1903	Device Explantation	A070504	Premature Discharge of Battery	10/12/2020	10/12/2020
					CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	10/12/2020	10/12/2020
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	10/12/2020	10/12/2020
							F12	Serious Injury/ Illness/ Impairment	A070504	Premature Discharge of Battery	10/12/2020	10/12/2020
							F1903	Device Explantation	A070504	Premature Discharge of Battery	10/12/2020	10/12/2020
					CT233	Surgical	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	10/12/2020	10/12/2020
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	10/12/2020	10/12/2020
							F12	Serious Injury/ Illness/ Impairment	A070504	Premature Discharge of Battery	10/12/2020	10/12/2020
							F1903	Device Explantation	A070504	Premature Discharge of Battery	10/12/2020	10/12/2020
					CT2406	Implantable	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	10/12/2020	10/12/2020
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	10/12/2020	10/12/2020
							F12	Serious Injury/ Illness/ Impairment	A070504	Premature Discharge of Battery	10/12/2020	10/12/2020
							F1903	Device Explantation	A070504	Premature Discharge of Battery	10/12/2020	10/12/2020
					CT2407	Active implantable	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	10/12/2020	10/12/2020
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	10/12/2020	10/12/2020
							F12	Serious Injury/ Illness/ Impairment	A070504	Premature Discharge of Battery	10/12/2020	10/12/2020
							F1903	Device Explantation	A070504	Premature Discharge of Battery	10/12/2020	10/12/2020

					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	10/12/2020	10/12/2020
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	10/12/2020	10/12/2020
							F12	Serious Injury/ Illness/ Impairment	A070504	Premature Discharge of Battery	10/12/2020	10/12/2020
							F1903	Device Explantation	A070504	Premature Discharge of Battery	10/12/2020	10/12/2020
					CT2727	Active implant devices/ancillaries	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	10/12/2020	10/12/2020
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	10/12/2020	10/12/2020
							F12	Serious Injury/ Illness/ Impairment	A070504	Premature Discharge of Battery	10/12/2020	10/12/2020
							F1903	Device Explantation	A070504	Premature Discharge of Battery	10/12/2020	10/12/2020
					CT2750	Implantable cardiac electrical stimulators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	10/12/2020	10/12/2020
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	10/12/2020	10/12/2020
							F12	Serious Injury/ Illness/ Impairment	A070504	Premature Discharge of Battery	10/12/2020	10/12/2020
							F1903	Device Explantation	A070504	Premature Discharge of Battery	10/12/2020	10/12/2020
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	10/12/2020	10/12/2020
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	10/12/2020	10/12/2020
							F12	Serious Injury/ Illness/ Impairment	A070504	Premature Discharge of Battery	10/12/2020	10/12/2020
							F1903	Device Explantation	A070504	Premature Discharge of Battery	10/12/2020	10/12/2020
					CT2985	Active	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	10/12/2020	10/12/2020
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	10/12/2020	10/12/2020
							F12	Serious Injury/ Illness/ Impairment	A070504	Premature Discharge of Battery	10/12/2020	10/12/2020
							F1903	Device Explantation	A070504	Premature Discharge of Battery	10/12/2020	10/12/2020
					CT3001	Implantable metal	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	10/12/2020	10/12/2020
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	10/12/2020	10/12/2020
							F12	Serious Injury/ Illness/ Impairment	A070504	Premature Discharge of Battery	10/12/2020	10/12/2020
							F1903	Device Explantation	A070504	Premature Discharge of Battery	10/12/2020	10/12/2020
					CT3002	Implantable synthetic polymer	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	10/12/2020	10/12/2020
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	10/12/2020	10/12/2020
							F12	Serious Injury/ Illness/ Impairment	A070504	Premature Discharge of Battery	10/12/2020	10/12/2020
							F1903	Device Explantation	A070504	Premature Discharge of Battery	10/12/2020	10/12/2020
					CT3006	Implantable inorganic material	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	10/12/2020	10/12/2020
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	10/12/2020	10/12/2020
							F12	Serious Injury/ Illness/ Impairment	A070504	Premature Discharge of Battery	10/12/2020	10/12/2020
							F1903	Device Explantation	A070504	Premature Discharge of Battery	10/12/2020	10/12/2020
					CT314	Electrically powered	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	10/12/2020	10/12/2020
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	10/12/2020	10/12/2020
							F12	Serious Injury/ Illness/ Impairment	A070504	Premature Discharge of Battery	10/12/2020	10/12/2020
							F1903	Device Explantation	A070504	Premature Discharge of Battery	10/12/2020	10/12/2020
					CT321	Long-term surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	10/12/2020	10/12/2020
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	10/12/2020	10/12/2020
							F12	Serious Injury/ Illness/ Impairment	A070504	Premature Discharge of Battery	10/12/2020	10/12/2020
							F1903	Device Explantation	A070504	Premature Discharge of Battery	10/12/2020	10/12/2020
					CT449	Generators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	10/12/2020	10/12/2020
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	10/12/2020	10/12/2020
							F12	Serious Injury/ Illness/ Impairment	A070504	Premature Discharge of Battery	10/12/2020	10/12/2020
							F1903	Device Explantation	A070504	Premature Discharge of Battery	10/12/2020	10/12/2020
					CT981	Single use	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	10/12/2020	10/12/2020
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	10/12/2020	10/12/2020
							F12	Serious Injury/ Illness/ Impairment	A070504	Premature Discharge of Battery	10/12/2020	10/12/2020
							F1903	Device Explantation	A070504	Premature Discharge of Battery	10/12/2020	10/12/2020
					CT983	Surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	10/12/2020	10/12/2020
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	10/12/2020	10/12/2020
							F12	Serious Injury/ Illness/ Impairment	A070504	Premature Discharge of Battery	10/12/2020	10/12/2020
							F1903	Device Explantation	A070504	Premature Discharge of Battery	10/12/2020	10/12/2020
10261	2020	All other reportable incidents	47263	Cardiac resynchronization therapy implantable pacemaker	CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070909	Oversensing		
									A0712	Pacing Problem		
									A090801	Signal Artifact/Noise		
							F26	No Health Consequences or Impact	A070909	Oversensing		
									A0712	Pacing Problem		
									A090801	Signal Artifact/Noise		
					CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A070909	Oversensing		
									A0712	Pacing Problem		
									A090801	Signal Artifact/Noise		
							F26	No Health Consequences or Impact	A070909	Oversensing		
									A0712	Pacing Problem		
									A090801	Signal Artifact/Noise		
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A070909	Oversensing		
									A0712	Pacing Problem		
									A090801	Signal Artifact/Noise		
							F26	No Health Consequences or Impact	A070909	Oversensing		
									A0712	Pacing Problem		
									A090801	Signal Artifact/Noise		
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A070909	Oversensing		
									A0712	Pacing Problem		
									A090801	Signal Artifact/Noise		
							F26	No Health Consequences or Impact	A070909	Oversensing		
									A0712	Pacing Problem		
									A090801	Signal Artifact/Noise		
10262	2020	All other reportable incidents	35852	Single-chamber implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A072201	High Impedance		
							F26	No Health Consequences or Impact	A072201	High Impedance		
10263	2020	All other reportable incidents	35223	Endocardial/interventricular septal pacing lead	CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070101	Failure to Capture	28/10/2020	28/10/2020
									A070909	Oversensing	28/10/2020	28/10/2020
									A0712	Pacing Problem	28/10/2020	28/10/2020
									A072201	High Impedance	28/10/2020	28/10/2020
									A12	Connection Problem	28/10/2020	28/10/2020
							F1901	Additional Surgery	A070101	Failure to Capture	28/10/2020	28/10/2020
									A070909	Oversensing	28/10/2020	28/10/2020
									A0712	Pacing Problem	28/10/2020	28/10/2020
									A072201	High Impedance	28/10/2020	28/10/2020
									A12	Connection Problem	28/10/2020	28/10/2020

						F1905	Device Revision or Replacement	A070101	Failure to Capture	28/10/2020	28/10/2020
								A070909	Oversensing	28/10/2020	28/10/2020
								A0712	Pacing Problem	28/10/2020	28/10/2020
								A072201	High Impedance	28/10/2020	28/10/2020
								A12	Connection Problem	28/10/2020	28/10/2020
10264	2020	All other reportable incidents	47267	Single-chamber implantable pacemaker, rate-responsive	CT141	Electrodes/Leads and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	03/11/2020	03/11/2020
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	03/11/2020	03/11/2020
							F12	Serious Injury/ Illness/ Impairment	A070504	Premature Discharge of Battery	03/11/2020	03/11/2020
							F19	Surgical Intervention	A070504	Premature Discharge of Battery	03/11/2020	03/11/2020
					CT145	Cardiovascular devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	03/11/2020	03/11/2020
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	03/11/2020	03/11/2020
							F12	Serious Injury/ Illness/ Impairment	A070504	Premature Discharge of Battery	03/11/2020	03/11/2020
							F19	Surgical Intervention	A070504	Premature Discharge of Battery	03/11/2020	03/11/2020
					CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	03/11/2020	03/11/2020
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	03/11/2020	03/11/2020
							F12	Serious Injury/ Illness/ Impairment	A070504	Premature Discharge of Battery	03/11/2020	03/11/2020
							F19	Surgical Intervention	A070504	Premature Discharge of Battery	03/11/2020	03/11/2020
					CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	03/11/2020	03/11/2020
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	03/11/2020	03/11/2020
							F12	Serious Injury/ Illness/ Impairment	A070504	Premature Discharge of Battery	03/11/2020	03/11/2020
							F19	Surgical Intervention	A070504	Premature Discharge of Battery	03/11/2020	03/11/2020
					CT233	Surgical	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	03/11/2020	03/11/2020
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	03/11/2020	03/11/2020
							F12	Serious Injury/ Illness/ Impairment	A070504	Premature Discharge of Battery	03/11/2020	03/11/2020
							F19	Surgical Intervention	A070504	Premature Discharge of Battery	03/11/2020	03/11/2020
					CT2406	Implantable	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	03/11/2020	03/11/2020
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	03/11/2020	03/11/2020
							F12	Serious Injury/ Illness/ Impairment	A070504	Premature Discharge of Battery	03/11/2020	03/11/2020
							F19	Surgical Intervention	A070504	Premature Discharge of Battery	03/11/2020	03/11/2020
					CT2407	Active implantable	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	03/11/2020	03/11/2020
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	03/11/2020	03/11/2020
							F12	Serious Injury/ Illness/ Impairment	A070504	Premature Discharge of Battery	03/11/2020	03/11/2020
							F19	Surgical Intervention	A070504	Premature Discharge of Battery	03/11/2020	03/11/2020
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	03/11/2020	03/11/2020
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	03/11/2020	03/11/2020
							F12	Serious Injury/ Illness/ Impairment	A070504	Premature Discharge of Battery	03/11/2020	03/11/2020
							F19	Surgical Intervention	A070504	Premature Discharge of Battery	03/11/2020	03/11/2020
					CT2727	Active implant devices/ancillaries	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	03/11/2020	03/11/2020
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	03/11/2020	03/11/2020
							F12	Serious Injury/ Illness/ Impairment	A070504	Premature Discharge of Battery	03/11/2020	03/11/2020
							F19	Surgical Intervention	A070504	Premature Discharge of Battery	03/11/2020	03/11/2020
					CT2750	Implantable cardiac electrical stimulators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	03/11/2020	03/11/2020
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	03/11/2020	03/11/2020
							F12	Serious Injury/ Illness/ Impairment	A070504	Premature Discharge of Battery	03/11/2020	03/11/2020
							F19	Surgical Intervention	A070504	Premature Discharge of Battery	03/11/2020	03/11/2020
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	03/11/2020	03/11/2020
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	03/11/2020	03/11/2020
							F12	Serious Injury/ Illness/ Impairment	A070504	Premature Discharge of Battery	03/11/2020	03/11/2020
							F19	Surgical Intervention	A070504	Premature Discharge of Battery	03/11/2020	03/11/2020
					CT2985	Active	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	03/11/2020	03/11/2020
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	03/11/2020	03/11/2020
							F12	Serious Injury/ Illness/ Impairment	A070504	Premature Discharge of Battery	03/11/2020	03/11/2020
							F19	Surgical Intervention	A070504	Premature Discharge of Battery	03/11/2020	03/11/2020
					CT3001	Implantable metal	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	03/11/2020	03/11/2020
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	03/11/2020	03/11/2020
							F12	Serious Injury/ Illness/ Impairment	A070504	Premature Discharge of Battery	03/11/2020	03/11/2020
							F19	Surgical Intervention	A070504	Premature Discharge of Battery	03/11/2020	03/11/2020
					CT3002	Implantable synthetic polymer	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	03/11/2020	03/11/2020
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	03/11/2020	03/11/2020
							F12	Serious Injury/ Illness/ Impairment	A070504	Premature Discharge of Battery	03/11/2020	03/11/2020
							F19	Surgical Intervention	A070504	Premature Discharge of Battery	03/11/2020	03/11/2020
					CT3006	Implantable inorganic material	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	03/11/2020	03/11/2020
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	03/11/2020	03/11/2020
							F12	Serious Injury/ Illness/ Impairment	A070504	Premature Discharge of Battery	03/11/2020	03/11/2020
							F19	Surgical Intervention	A070504	Premature Discharge of Battery	03/11/2020	03/11/2020
					CT314	Electrically powered	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	03/11/2020	03/11/2020
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	03/11/2020	03/11/2020
							F12	Serious Injury/ Illness/ Impairment	A070504	Premature Discharge of Battery	03/11/2020	03/11/2020
							F19	Surgical Intervention	A070504	Premature Discharge of Battery	03/11/2020	03/11/2020
					CT321	Long-term surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	03/11/2020	03/11/2020
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	03/11/2020	03/11/2020
							F12	Serious Injury/ Illness/ Impairment	A070504	Premature Discharge of Battery	03/11/2020	03/11/2020
							F19	Surgical Intervention	A070504	Premature Discharge of Battery	03/11/2020	03/11/2020
					CT449	Generators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	03/11/2020	03/11/2020
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	03/11/2020	03/11/2020
							F12	Serious Injury/ Illness/ Impairment	A070504	Premature Discharge of Battery	03/11/2020	03/11/2020
							F19	Surgical Intervention	A070504	Premature Discharge of Battery	03/11/2020	03/11/2020
					CT981	Single use	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	03/11/2020	03/11/2020
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	03/11/2020	03/11/2020
							F12	Serious Injury/ Illness/ Impairment	A070504	Premature Discharge of Battery	03/11/2020	03/11/2020
							F19	Surgical Intervention	A070504	Premature Discharge of Battery	03/11/2020	03/11/2020
					CT983	Surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	03/11/2020	03/11/2020
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	03/11/2020	03/11/2020
							F12	Serious Injury/ Illness/ Impairment	A070504	Premature Discharge of Battery	03/11/2020	03/11/2020
							F19	Surgical Intervention	A070504	Premature Discharge of Battery	03/11/2020	03/11/2020
10265	2020	All other reportable incidents	35852	Single-chamber implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	15/12/2020	15/12/2020
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	15/12/2020	15/12/2020
							F12	Serious Injury/ Illness/ Impairment	A070504	Premature Discharge of Battery	15/12/2020	15/12/2020

							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	15/12/2020	15/12/2020
10266	2020	All other reportable incidents	47265	Dual-chamber implantable pacemaker, rate-responsive	CT141	Electrodes/Leads and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT145	Cardiovascular devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT233	Surgical	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT2406	Implantable	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT2407	Active implantable	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT2727	Active implant devices/ancillaries	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT2750	Implantable cardiac electrical stimulators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT2985	Active	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT3001	Implantable metal	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT3002	Implantable synthetic polymer	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT3006	Implantable inorganic material	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT314	Electrically powered	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT321	Long-term surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT449	Generators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT981	Single use	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT983	Surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
10267	2020	All other reportable incidents	47265	Dual-chamber implantable pacemaker, rate-responsive	CT141	Electrodes/Leads and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT145	Cardiovascular devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT233	Surgical	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT2406	Implantable	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT2407	Active implantable	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT2727	Active implant devices/ancillaries	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT2750	Implantable cardiac electrical stimulators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT2985	Active	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT3001	Implantable metal	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT3002	Implantable synthetic polymer	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT3006	Implantable inorganic material	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT314	Electrically powered	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT321	Long-term surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT449	Generators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT981	Single use	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT983	Surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
10268	2020	All other reportable incidents	47265	Dual-chamber implantable pacemaker, rate-responsive	CT141	Electrodes/Leads and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	16/10/2020	16/10/2020
							F1901	Additional Surgery	A070504	Premature Discharge of Battery	16/10/2020	16/10/2020
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	16/10/2020	16/10/2020
					CT145	Cardiovascular devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	16/10/2020	16/10/2020
							F1901	Additional Surgery	A070504	Premature Discharge of Battery	16/10/2020	16/10/2020
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	16/10/2020	16/10/2020
					CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	16/10/2020	16/10/2020

							F1901	Additional Surgery	A070504	Premature Discharge of Battery	16/10/2020	16/10/2020
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	16/10/2020	16/10/2020
					CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	16/10/2020	16/10/2020
							F1901	Additional Surgery	A070504	Premature Discharge of Battery	16/10/2020	16/10/2020
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	16/10/2020	16/10/2020
					CT233	Surgical	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	16/10/2020	16/10/2020
							F1901	Additional Surgery	A070504	Premature Discharge of Battery	16/10/2020	16/10/2020
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	16/10/2020	16/10/2020
					CT2406	Implantable	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	16/10/2020	16/10/2020
							F1901	Additional Surgery	A070504	Premature Discharge of Battery	16/10/2020	16/10/2020
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	16/10/2020	16/10/2020
					CT2407	Active implantable	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	16/10/2020	16/10/2020
							F1901	Additional Surgery	A070504	Premature Discharge of Battery	16/10/2020	16/10/2020
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	16/10/2020	16/10/2020
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	16/10/2020	16/10/2020
							F1901	Additional Surgery	A070504	Premature Discharge of Battery	16/10/2020	16/10/2020
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	16/10/2020	16/10/2020
					CT2727	Active implant devices/ancillaries	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	16/10/2020	16/10/2020
							F1901	Additional Surgery	A070504	Premature Discharge of Battery	16/10/2020	16/10/2020
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	16/10/2020	16/10/2020
					CT2750	Implantable cardiac electrical stimulators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	16/10/2020	16/10/2020
							F1901	Additional Surgery	A070504	Premature Discharge of Battery	16/10/2020	16/10/2020
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	16/10/2020	16/10/2020
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	16/10/2020	16/10/2020
							F1901	Additional Surgery	A070504	Premature Discharge of Battery	16/10/2020	16/10/2020
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	16/10/2020	16/10/2020
					CT2985	Active	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	16/10/2020	16/10/2020
							F1901	Additional Surgery	A070504	Premature Discharge of Battery	16/10/2020	16/10/2020
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	16/10/2020	16/10/2020
					CT3001	Implantable metal	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	16/10/2020	16/10/2020
							F1901	Additional Surgery	A070504	Premature Discharge of Battery	16/10/2020	16/10/2020
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	16/10/2020	16/10/2020
					CT3002	Implantable synthetic polymer	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	16/10/2020	16/10/2020
							F1901	Additional Surgery	A070504	Premature Discharge of Battery	16/10/2020	16/10/2020
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	16/10/2020	16/10/2020
					CT3006	Implantable inorganic material	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	16/10/2020	16/10/2020
							F1901	Additional Surgery	A070504	Premature Discharge of Battery	16/10/2020	16/10/2020
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	16/10/2020	16/10/2020
					CT314	Electrically powered	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	16/10/2020	16/10/2020
							F1901	Additional Surgery	A070504	Premature Discharge of Battery	16/10/2020	16/10/2020
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	16/10/2020	16/10/2020
					CT321	Long-term surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	16/10/2020	16/10/2020
							F1901	Additional Surgery	A070504	Premature Discharge of Battery	16/10/2020	16/10/2020
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	16/10/2020	16/10/2020
					CT449	Generators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	16/10/2020	16/10/2020
							F1901	Additional Surgery	A070504	Premature Discharge of Battery	16/10/2020	16/10/2020
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	16/10/2020	16/10/2020
					CT981	Single use	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	16/10/2020	16/10/2020
							F1901	Additional Surgery	A070504	Premature Discharge of Battery	16/10/2020	16/10/2020
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	16/10/2020	16/10/2020
					CT983	Surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	16/10/2020	16/10/2020
							F1901	Additional Surgery	A070504	Premature Discharge of Battery	16/10/2020	16/10/2020
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	16/10/2020	16/10/2020
10269	2020	All other reportable incidents	60789	Intracardiac pacemaker	CT130	Catheters and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A071102	Failure to Interrogate		
							F1901	Additional Surgery	A071102	Failure to Interrogate		
							F1905	Device Revision or Replacement	A071102	Failure to Interrogate		
					CT145	Cardiovascular devices	E2403	No Clinical Signs, Symptoms or Conditions	A071102	Failure to Interrogate		
							F1901	Additional Surgery	A071102	Failure to Interrogate		
							F1905	Device Revision or Replacement	A071102	Failure to Interrogate		
					CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A071102	Failure to Interrogate		
							F1901	Additional Surgery	A071102	Failure to Interrogate		
							F1905	Device Revision or Replacement	A071102	Failure to Interrogate		
					CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A071102	Failure to Interrogate		
							F1901	Additional Surgery	A071102	Failure to Interrogate		
							F1905	Device Revision or Replacement	A071102	Failure to Interrogate		
					CT233	Surgical	E2403	No Clinical Signs, Symptoms or Conditions	A071102	Failure to Interrogate		
							F1901	Additional Surgery	A071102	Failure to Interrogate		
							F1905	Device Revision or Replacement	A071102	Failure to Interrogate		
					CT2406	Implantable	E2403	No Clinical Signs, Symptoms or Conditions	A071102	Failure to Interrogate		
							F1901	Additional Surgery	A071102	Failure to Interrogate		
							F1905	Device Revision or Replacement	A071102	Failure to Interrogate		
					CT2407	Active implantable	E2403	No Clinical Signs, Symptoms or Conditions	A071102	Failure to Interrogate		
							F1901	Additional Surgery	A071102	Failure to Interrogate		
							F1905	Device Revision or Replacement	A071102	Failure to Interrogate		
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A071102	Failure to Interrogate		
							F1901	Additional Surgery	A071102	Failure to Interrogate		
							F1905	Device Revision or Replacement	A071102	Failure to Interrogate		
					CT2727	Active implant devices/ancillaries	E2403	No Clinical Signs, Symptoms or Conditions	A071102	Failure to Interrogate		
							F1901	Additional Surgery	A071102	Failure to Interrogate		
							F1905	Device Revision or Replacement	A071102	Failure to Interrogate		
					CT2750	Implantable cardiac electrical stimulators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A071102	Failure to Interrogate		
							F1901	Additional Surgery	A071102	Failure to Interrogate		
							F1905	Device Revision or Replacement	A071102	Failure to Interrogate		
					CT2818	Catheterization/Cannulation/Fluid path devices	E2403	No Clinical Signs, Symptoms or Conditions	A071102	Failure to Interrogate		
							F1901	Additional Surgery	A071102	Failure to Interrogate		
							F1905	Device Revision or Replacement	A071102	Failure to Interrogate		
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A071102	Failure to Interrogate		
							F1901	Additional Surgery	A071102	Failure to Interrogate		

							F1905	Device Revision or Replacement	A071102	Failure to Interrogate			
					CT2985	Active	E2403	No Clinical Signs, Symptoms or Conditions	A071102	Failure to Interrogate			
							F1901	Additional Surgery	A071102	Failure to Interrogate			
							F1905	Device Revision or Replacement	A071102	Failure to Interrogate			
					CT3001	Implantable metal	E2403	No Clinical Signs, Symptoms or Conditions	A071102	Failure to Interrogate			
							F1901	Additional Surgery	A071102	Failure to Interrogate			
							F1905	Device Revision or Replacement	A071102	Failure to Interrogate			
					CT3002	Implantable synthetic polymer	E2403	No Clinical Signs, Symptoms or Conditions	A071102	Failure to Interrogate			
							F1901	Additional Surgery	A071102	Failure to Interrogate			
							F1905	Device Revision or Replacement	A071102	Failure to Interrogate			
					CT3006	Implantable inorganic material	E2403	No Clinical Signs, Symptoms or Conditions	A071102	Failure to Interrogate			
							F1901	Additional Surgery	A071102	Failure to Interrogate			
							F1905	Device Revision or Replacement	A071102	Failure to Interrogate			
					CT314	Electrically powered	E2403	No Clinical Signs, Symptoms or Conditions	A071102	Failure to Interrogate			
							F1901	Additional Surgery	A071102	Failure to Interrogate			
							F1905	Device Revision or Replacement	A071102	Failure to Interrogate			
					CT321	Long-term surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A071102	Failure to Interrogate			
							F1901	Additional Surgery	A071102	Failure to Interrogate			
							F1905	Device Revision or Replacement	A071102	Failure to Interrogate			
					CT981	Single use	E2403	No Clinical Signs, Symptoms or Conditions	A071102	Failure to Interrogate			
							F1901	Additional Surgery	A071102	Failure to Interrogate			
							F1905	Device Revision or Replacement	A071102	Failure to Interrogate			
					CT983	Surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A071102	Failure to Interrogate			
							F1901	Additional Surgery	A071102	Failure to Interrogate			
							F1905	Device Revision or Replacement	A071102	Failure to Interrogate			
10270	2020	All other reportable incidents	47263	Cardiac resynchronization therapy implantable pacemaker	CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070102	High Capture Threshold	30/10/2020	30/10/2020	
									A070504	Premature Discharge of Battery	30/10/2020	30/10/2020	
									A090801	Signal Artifact/Noise	30/10/2020	30/10/2020	
									A13	Communication or Transmission Problem	30/10/2020	30/10/2020	
							F19	Surgical Intervention	A070102	High Capture Threshold	30/10/2020	30/10/2020	
									A070504	Premature Discharge of Battery	30/10/2020	30/10/2020	
									A090801	Signal Artifact/Noise	30/10/2020	30/10/2020	
									A13	Communication or Transmission Problem	30/10/2020	30/10/2020	
					CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A070102	High Capture Threshold	30/10/2020	30/10/2020	
									A070504	Premature Discharge of Battery	30/10/2020	30/10/2020	
									A090801	Signal Artifact/Noise	30/10/2020	30/10/2020	
									A13	Communication or Transmission Problem	30/10/2020	30/10/2020	
							F19	Surgical Intervention	A070102	High Capture Threshold	30/10/2020	30/10/2020	
									A070504	Premature Discharge of Battery	30/10/2020	30/10/2020	
									A090801	Signal Artifact/Noise	30/10/2020	30/10/2020	
									A13	Communication or Transmission Problem	30/10/2020	30/10/2020	
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A070102	High Capture Threshold	30/10/2020	30/10/2020	
									A070504	Premature Discharge of Battery	30/10/2020	30/10/2020	
									A090801	Signal Artifact/Noise	30/10/2020	30/10/2020	
									A13	Communication or Transmission Problem	30/10/2020	30/10/2020	
							F19	Surgical Intervention	A070102	High Capture Threshold	30/10/2020	30/10/2020	
									A070504	Premature Discharge of Battery	30/10/2020	30/10/2020	
									A090801	Signal Artifact/Noise	30/10/2020	30/10/2020	
									A13	Communication or Transmission Problem	30/10/2020	30/10/2020	
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A070102	High Capture Threshold	30/10/2020	30/10/2020	
									A070504	Premature Discharge of Battery	30/10/2020	30/10/2020	
									A090801	Signal Artifact/Noise	30/10/2020	30/10/2020	
									A13	Communication or Transmission Problem	30/10/2020	30/10/2020	
							F19	Surgical Intervention	A070102	High Capture Threshold	30/10/2020	30/10/2020	
									A070504	Premature Discharge of Battery	30/10/2020	30/10/2020	
									A090801	Signal Artifact/Noise	30/10/2020	30/10/2020	
									A13	Communication or Transmission Problem	30/10/2020	30/10/2020	
10271	2020	All other reportable incidents	37265	Dual-chamber implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	17/11/2020	17/11/2020	
							F19	Surgical Intervention	A070504	Premature Discharge of Battery	17/11/2020	17/11/2020	
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	17/11/2020	17/11/2020	
10272	2020	All other reportable incidents	35852	Single-chamber implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A072201	High Impedance			
							F26	No Health Consequences or Impact	A072201	High Impedance			
10273	2020	All other reportable incidents	36241	Pacemaker/defibrillator lead	CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A072201	High Impedance			
							F26	No Health Consequences or Impact	A072201	High Impedance			
10274	2020	All other reportable incidents	47270	Cardiac resynchronization therapy implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	13/11/2020	13/11/2020	
							F1901	Additional Surgery	A070504	Premature Discharge of Battery	13/11/2020	13/11/2020	
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	13/11/2020	13/11/2020	
10275	2020	All other reportable incidents	47265	Dual-chamber implantable pacemaker, rate-responsive	CT141	Electrodes/Leads and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery			
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery			
					CT145	Cardiovascular devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery			
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery			
					CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery			
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery			
					CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery			
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery			
					CT233	Surgical	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery			
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery			
					CT2406	Implantable	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery			
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery			
					CT2407	Active implantable	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery			
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery			
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery			
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery			
					CT2727	Active implant devices/ancillaries	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery			
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery			
					CT2750	Implantable cardiac electrical stimulators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery			
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery			
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery			

						F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery			
				CT2985	Active	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery			
						F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery			
				CT3001	Implantable metal	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery			
						F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery			
				CT3002	Implantable synthetic polymer	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery			
						F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery			
				CT3006	Implantable inorganic material	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery			
						F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery			
				CT314	Electrically powered	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery			
						F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery			
				CT321	Long-term surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery			
						F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery			
				CT449	Generators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery			
						F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery			
				CT981	Single use	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery			
						F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery			
				CT983	Surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery			
						F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery			
10276	2020	All other reportable incidents	36241	Pacemaker/defibrillator lead	CT1684	Cardiac pacemakers and associated devices	E2114	Perforation	A04	Material Integrity Problem	26/10/2020	26/10/2020
							F08	Hospitalization or Prolonged Hospitalization	A04	Material Integrity Problem	26/10/2020	26/10/2020
							F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A04	Material Integrity Problem	26/10/2020	26/10/2020
							F12	Serious Injury/ Illness/ Impairment	A04	Material Integrity Problem	26/10/2020	26/10/2020
							F1905	Device Revision or Replacement	A04	Material Integrity Problem	26/10/2020	26/10/2020
10277	2020	All other reportable incidents	47267	Single-chamber implantable pacemaker, rate-responsive	CT141	Electrodes/Leads and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	02/12/2020	02/12/2020
							F19	Surgical Intervention	A070504	Premature Discharge of Battery	02/12/2020	02/12/2020
					CT145	Cardiovascular devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	02/12/2020	02/12/2020
							F19	Surgical Intervention	A070504	Premature Discharge of Battery	02/12/2020	02/12/2020
					CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	02/12/2020	02/12/2020
							F19	Surgical Intervention	A070504	Premature Discharge of Battery	02/12/2020	02/12/2020
					CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	02/12/2020	02/12/2020
							F19	Surgical Intervention	A070504	Premature Discharge of Battery	02/12/2020	02/12/2020
					CT233	Surgical	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	02/12/2020	02/12/2020
							F19	Surgical Intervention	A070504	Premature Discharge of Battery	02/12/2020	02/12/2020
					CT2406	Implantable	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	02/12/2020	02/12/2020
							F19	Surgical Intervention	A070504	Premature Discharge of Battery	02/12/2020	02/12/2020
					CT2407	Active implantable	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	02/12/2020	02/12/2020
							F19	Surgical Intervention	A070504	Premature Discharge of Battery	02/12/2020	02/12/2020
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	02/12/2020	02/12/2020
							F19	Surgical Intervention	A070504	Premature Discharge of Battery	02/12/2020	02/12/2020
					CT2727	Active implant devices/ancillaries	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	02/12/2020	02/12/2020
							F19	Surgical Intervention	A070504	Premature Discharge of Battery	02/12/2020	02/12/2020
					CT2750	Implantable cardiac electrical stimulators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	02/12/2020	02/12/2020
							F19	Surgical Intervention	A070504	Premature Discharge of Battery	02/12/2020	02/12/2020
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	02/12/2020	02/12/2020
							F19	Surgical Intervention	A070504	Premature Discharge of Battery	02/12/2020	02/12/2020
					CT2985	Active	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	02/12/2020	02/12/2020
							F19	Surgical Intervention	A070504	Premature Discharge of Battery	02/12/2020	02/12/2020
					CT3001	Implantable metal	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	02/12/2020	02/12/2020
							F19	Surgical Intervention	A070504	Premature Discharge of Battery	02/12/2020	02/12/2020
					CT3002	Implantable synthetic polymer	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	02/12/2020	02/12/2020
							F19	Surgical Intervention	A070504	Premature Discharge of Battery	02/12/2020	02/12/2020
					CT3006	Implantable inorganic material	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	02/12/2020	02/12/2020
							F19	Surgical Intervention	A070504	Premature Discharge of Battery	02/12/2020	02/12/2020
					CT314	Electrically powered	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	02/12/2020	02/12/2020
							F19	Surgical Intervention	A070504	Premature Discharge of Battery	02/12/2020	02/12/2020
					CT321	Long-term surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	02/12/2020	02/12/2020
							F19	Surgical Intervention	A070504	Premature Discharge of Battery	02/12/2020	02/12/2020
					CT449	Generators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	02/12/2020	02/12/2020
							F19	Surgical Intervention	A070504	Premature Discharge of Battery	02/12/2020	02/12/2020
					CT981	Single use	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	02/12/2020	02/12/2020
							F19	Surgical Intervention	A070504	Premature Discharge of Battery	02/12/2020	02/12/2020
					CT983	Surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	02/12/2020	02/12/2020
							F19	Surgical Intervention	A070504	Premature Discharge of Battery	02/12/2020	02/12/2020
10278	2020	All other reportable incidents	35852	Single-chamber implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	27/11/2020	27/11/2020
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	27/11/2020	27/11/2020
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	27/11/2020	27/11/2020
10279	2020	All other reportable incidents	35852	Single-chamber implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
10280	2020	All other reportable incidents	36241	Pacemaker/defibrillator lead	CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A04	Material Integrity Problem		
									A070909	Oversensing		
									A072201	High Impedance		
									A090801	Signal Artifact/Noise		
							F26	No Health Consequences or Impact	A04	Material Integrity Problem		
									A070909	Oversensing		
									A072201	High Impedance		
									A090801	Signal Artifact/Noise		
10281	2020	All other reportable incidents	36241	Pacemaker/defibrillator lead	CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A04	Material Integrity Problem		
									A070909	Oversensing		
									A0712	Pacing Problem		
									A090801	Signal Artifact/Noise		
							F23	Unexpected Medical Intervention	A04	Material Integrity Problem		
									A070909	Oversensing		
									A0712	Pacing Problem		
									A090801	Signal Artifact/Noise		
10282	2020	All other reportable incidents	47265	Dual-chamber implantable pacemaker, rate-responsive	CT141	Electrodes/Leads and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070909	Oversensing		
									A0712	Pacing Problem		
									A072201	High Impedance		

									A090801	Signal Artifact/Noise		
						F23	Unexpected Medical Intervention	A070909	Oversensing			
								A0712	Pacing Problem			
								A072201	High Impedance			
								A090801	Signal Artifact/Noise			
					CT145	Cardiovascular devices	E2403	No Clinical Signs, Symptoms or Conditions	A070909	Oversensing		
									A0712	Pacing Problem		
									A072201	High Impedance		
									A090801	Signal Artifact/Noise		
							F23	Unexpected Medical Intervention	A070909	Oversensing		
									A0712	Pacing Problem		
									A072201	High Impedance		
									A090801	Signal Artifact/Noise		
					CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070909	Oversensing		
									A0712	Pacing Problem		
									A072201	High Impedance		
									A090801	Signal Artifact/Noise		
							F23	Unexpected Medical Intervention	A070909	Oversensing		
									A0712	Pacing Problem		
									A072201	High Impedance		
									A090801	Signal Artifact/Noise		
					CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A070909	Oversensing		
									A0712	Pacing Problem		
									A072201	High Impedance		
									A090801	Signal Artifact/Noise		
							F23	Unexpected Medical Intervention	A070909	Oversensing		
									A0712	Pacing Problem		
									A072201	High Impedance		
									A090801	Signal Artifact/Noise		
					CT233	Surgical	E2403	No Clinical Signs, Symptoms or Conditions	A070909	Oversensing		
									A0712	Pacing Problem		
									A072201	High Impedance		
									A090801	Signal Artifact/Noise		
							F23	Unexpected Medical Intervention	A070909	Oversensing		
									A0712	Pacing Problem		
									A072201	High Impedance		
									A090801	Signal Artifact/Noise		
					CT2406	Implantable	E2403	No Clinical Signs, Symptoms or Conditions	A070909	Oversensing		
									A0712	Pacing Problem		
									A072201	High Impedance		
									A090801	Signal Artifact/Noise		
							F23	Unexpected Medical Intervention	A070909	Oversensing		
									A0712	Pacing Problem		
									A072201	High Impedance		
									A090801	Signal Artifact/Noise		
					CT2407	Active implantable	E2403	No Clinical Signs, Symptoms or Conditions	A070909	Oversensing		
									A0712	Pacing Problem		
									A072201	High Impedance		
									A090801	Signal Artifact/Noise		
							F23	Unexpected Medical Intervention	A070909	Oversensing		
									A0712	Pacing Problem		
									A072201	High Impedance		
									A090801	Signal Artifact/Noise		
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A070909	Oversensing		
									A0712	Pacing Problem		
									A072201	High Impedance		
									A090801	Signal Artifact/Noise		
							F23	Unexpected Medical Intervention	A070909	Oversensing		
									A0712	Pacing Problem		
									A072201	High Impedance		
									A090801	Signal Artifact/Noise		
					CT2727	Active implant devices/ancillaries	E2403	No Clinical Signs, Symptoms or Conditions	A070909	Oversensing		
									A0712	Pacing Problem		
									A072201	High Impedance		
									A090801	Signal Artifact/Noise		
							F23	Unexpected Medical Intervention	A070909	Oversensing		
									A0712	Pacing Problem		
									A072201	High Impedance		
									A090801	Signal Artifact/Noise		
					CT2750	Implantable cardiac electrical stimulators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070909	Oversensing		
									A0712	Pacing Problem		
									A072201	High Impedance		
									A090801	Signal Artifact/Noise		
							F23	Unexpected Medical Intervention	A070909	Oversensing		
									A0712	Pacing Problem		
									A072201	High Impedance		
									A090801	Signal Artifact/Noise		
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A070909	Oversensing		
									A0712	Pacing Problem		
									A072201	High Impedance		
									A090801	Signal Artifact/Noise		
							F23	Unexpected Medical Intervention	A070909	Oversensing		
									A0712	Pacing Problem		
									A072201	High Impedance		
									A090801	Signal Artifact/Noise		
					CT2985	Active	E2403	No Clinical Signs, Symptoms or Conditions	A070909	Oversensing		
									A0712	Pacing Problem		
									A072201	High Impedance		

									A090801	Signal Artifact/Noise		
							F23	Unexpected Medical Intervention	A070909	Oversensing		
									A0712	Pacing Problem		
									A072201	High Impedance		
									A090801	Signal Artifact/Noise		
					CT3001	Implantable metal	E2403	No Clinical Signs, Symptoms or Conditions	A070909	Oversensing		
									A0712	Pacing Problem		
									A072201	High Impedance		
									A090801	Signal Artifact/Noise		
							F23	Unexpected Medical Intervention	A070909	Oversensing		
									A0712	Pacing Problem		
									A072201	High Impedance		
									A090801	Signal Artifact/Noise		
					CT3002	Implantable synthetic polymer	E2403	No Clinical Signs, Symptoms or Conditions	A070909	Oversensing		
									A0712	Pacing Problem		
									A072201	High Impedance		
									A090801	Signal Artifact/Noise		
							F23	Unexpected Medical Intervention	A070909	Oversensing		
									A0712	Pacing Problem		
									A072201	High Impedance		
									A090801	Signal Artifact/Noise		
					CT3006	Implantable inorganic material	E2403	No Clinical Signs, Symptoms or Conditions	A070909	Oversensing		
									A0712	Pacing Problem		
									A072201	High Impedance		
									A090801	Signal Artifact/Noise		
							F23	Unexpected Medical Intervention	A070909	Oversensing		
									A0712	Pacing Problem		
									A072201	High Impedance		
									A090801	Signal Artifact/Noise		
					CT314	Electrically powered	E2403	No Clinical Signs, Symptoms or Conditions	A070909	Oversensing		
									A0712	Pacing Problem		
									A072201	High Impedance		
									A090801	Signal Artifact/Noise		
							F23	Unexpected Medical Intervention	A070909	Oversensing		
									A0712	Pacing Problem		
									A072201	High Impedance		
									A090801	Signal Artifact/Noise		
					CT321	Long-term surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A070909	Oversensing		
									A0712	Pacing Problem		
									A072201	High Impedance		
									A090801	Signal Artifact/Noise		
							F23	Unexpected Medical Intervention	A070909	Oversensing		
									A0712	Pacing Problem		
									A072201	High Impedance		
									A090801	Signal Artifact/Noise		
					CT449	Generators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070909	Oversensing		
									A0712	Pacing Problem		
									A072201	High Impedance		
									A090801	Signal Artifact/Noise		
							F23	Unexpected Medical Intervention	A070909	Oversensing		
									A0712	Pacing Problem		
									A072201	High Impedance		
									A090801	Signal Artifact/Noise		
					CT981	Single use	E2403	No Clinical Signs, Symptoms or Conditions	A070909	Oversensing		
									A0712	Pacing Problem		
									A072201	High Impedance		
									A090801	Signal Artifact/Noise		
							F23	Unexpected Medical Intervention	A070909	Oversensing		
									A0712	Pacing Problem		
									A072201	High Impedance		
									A090801	Signal Artifact/Noise		
					CT983	Surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A070909	Oversensing		
									A0712	Pacing Problem		
									A072201	High Impedance		
									A090801	Signal Artifact/Noise		
							F23	Unexpected Medical Intervention	A070909	Oversensing		
									A0712	Pacing Problem		
									A072201	High Impedance		
									A090801	Signal Artifact/Noise		
10283	2020	All other reportable incidents	47270	Cardiac resynchronization therapy implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A070909	Oversensing		
									A072201	High Impedance		
									A12	Connection Problem		
							F26	No Health Consequences or Impact	A070909	Oversensing		
									A072201	High Impedance		
									A12	Connection Problem		
10284	2020	All other reportable incidents	47265	Dual-chamber implantable pacemaker, rate-responsive	CT141	Electrodes/Leads and associated devices	E2401	Insufficient Information	A070101	Failure to Capture		
							E2403	No Clinical Signs, Symptoms or Conditions	A070101	Failure to Capture		
							F23	Unexpected Medical Intervention	A070101	Failure to Capture		
					CT145	Cardiovascular devices	E2401	Insufficient Information	A070101	Failure to Capture		
							E2403	No Clinical Signs, Symptoms or Conditions	A070101	Failure to Capture		
							F23	Unexpected Medical Intervention	A070101	Failure to Capture		
					CT1684	Cardiac pacemakers and associated devices	E2401	Insufficient Information	A070101	Failure to Capture		
							E2403	No Clinical Signs, Symptoms or Conditions	A070101	Failure to Capture		
							F23	Unexpected Medical Intervention	A070101	Failure to Capture		
					CT1685	Cardiac pacemakers	E2401	Insufficient Information	A070101	Failure to Capture		
							E2403	No Clinical Signs, Symptoms or Conditions	A070101	Failure to Capture		
							F23	Unexpected Medical Intervention	A070101	Failure to Capture		
					CT233	Surgical	E2401	Insufficient Information	A070101	Failure to Capture		

							E2403	No Clinical Signs, Symptoms or Conditions	A070101	Failure to Capture		
							F23	Unexpected Medical Intervention	A070101	Failure to Capture		
					CT2406	Implantable	E2401	Insufficient Information	A070101	Failure to Capture		
							E2403	No Clinical Signs, Symptoms or Conditions	A070101	Failure to Capture		
							F23	Unexpected Medical Intervention	A070101	Failure to Capture		
					CT2407	Active implantable	E2401	Insufficient Information	A070101	Failure to Capture		
							E2403	No Clinical Signs, Symptoms or Conditions	A070101	Failure to Capture		
							F23	Unexpected Medical Intervention	A070101	Failure to Capture		
					CT257	Implantable cardiac pacemakers	E2401	Insufficient Information	A070101	Failure to Capture		
							E2403	No Clinical Signs, Symptoms or Conditions	A070101	Failure to Capture		
							F23	Unexpected Medical Intervention	A070101	Failure to Capture		
					CT2727	Active implant devices/ancillaries	E2401	Insufficient Information	A070101	Failure to Capture		
							E2403	No Clinical Signs, Symptoms or Conditions	A070101	Failure to Capture		
							F23	Unexpected Medical Intervention	A070101	Failure to Capture		
					CT2750	Implantable cardiac electrical stimulators and associated devices	E2401	Insufficient Information	A070101	Failure to Capture		
							E2403	No Clinical Signs, Symptoms or Conditions	A070101	Failure to Capture		
							F23	Unexpected Medical Intervention	A070101	Failure to Capture		
					CT2866	Implantable cardiac electrical stimulators	E2401	Insufficient Information	A070101	Failure to Capture		
							E2403	No Clinical Signs, Symptoms or Conditions	A070101	Failure to Capture		
							F23	Unexpected Medical Intervention	A070101	Failure to Capture		
					CT2985	Active	E2401	Insufficient Information	A070101	Failure to Capture		
							E2403	No Clinical Signs, Symptoms or Conditions	A070101	Failure to Capture		
							F23	Unexpected Medical Intervention	A070101	Failure to Capture		
					CT3001	Implantable metal	E2401	Insufficient Information	A070101	Failure to Capture		
							E2403	No Clinical Signs, Symptoms or Conditions	A070101	Failure to Capture		
							F23	Unexpected Medical Intervention	A070101	Failure to Capture		
					CT3002	Implantable synthetic polymer	E2401	Insufficient Information	A070101	Failure to Capture		
							E2403	No Clinical Signs, Symptoms or Conditions	A070101	Failure to Capture		
							F23	Unexpected Medical Intervention	A070101	Failure to Capture		
					CT3006	Implantable inorganic material	E2401	Insufficient Information	A070101	Failure to Capture		
							E2403	No Clinical Signs, Symptoms or Conditions	A070101	Failure to Capture		
							F23	Unexpected Medical Intervention	A070101	Failure to Capture		
					CT314	Electrically powered	E2401	Insufficient Information	A070101	Failure to Capture		
							E2403	No Clinical Signs, Symptoms or Conditions	A070101	Failure to Capture		
							F23	Unexpected Medical Intervention	A070101	Failure to Capture		
					CT321	Long-term surgical invasive	E2401	Insufficient Information	A070101	Failure to Capture		
							E2403	No Clinical Signs, Symptoms or Conditions	A070101	Failure to Capture		
							F23	Unexpected Medical Intervention	A070101	Failure to Capture		
					CT449	Generators and associated devices	E2401	Insufficient Information	A070101	Failure to Capture		
							E2403	No Clinical Signs, Symptoms or Conditions	A070101	Failure to Capture		
							F23	Unexpected Medical Intervention	A070101	Failure to Capture		
					CT981	Single use	E2401	Insufficient Information	A070101	Failure to Capture		
							E2403	No Clinical Signs, Symptoms or Conditions	A070101	Failure to Capture		
							F23	Unexpected Medical Intervention	A070101	Failure to Capture		
					CT983	Surgical invasive	E2401	Insufficient Information	A070101	Failure to Capture		
							E2403	No Clinical Signs, Symptoms or Conditions	A070101	Failure to Capture		
							F23	Unexpected Medical Intervention	A070101	Failure to Capture		
10285	2020	All other reportable incidents	47267	Single-chamber implantable pacemaker, rate-responsive	CT141	Electrodes/Leads and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	10/02/2022	10/02/2022
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	10/02/2022	10/02/2022
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	10/02/2022	10/02/2022
					CT145	Cardiovascular devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	10/02/2022	10/02/2022
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	10/02/2022	10/02/2022
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	10/02/2022	10/02/2022
					CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	10/02/2022	10/02/2022
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	10/02/2022	10/02/2022
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	10/02/2022	10/02/2022
					CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	10/02/2022	10/02/2022
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	10/02/2022	10/02/2022
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	10/02/2022	10/02/2022
					CT233	Surgical	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	10/02/2022	10/02/2022
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	10/02/2022	10/02/2022
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	10/02/2022	10/02/2022
					CT2406	Implantable	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	10/02/2022	10/02/2022
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	10/02/2022	10/02/2022
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	10/02/2022	10/02/2022
					CT2407	Active implantable	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	10/02/2022	10/02/2022
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	10/02/2022	10/02/2022
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	10/02/2022	10/02/2022
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	10/02/2022	10/02/2022
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	10/02/2022	10/02/2022
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	10/02/2022	10/02/2022
					CT2727	Active implant devices/ancillaries	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	10/02/2022	10/02/2022
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	10/02/2022	10/02/2022
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	10/02/2022	10/02/2022
					CT2750	Implantable cardiac electrical stimulators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	10/02/2022	10/02/2022
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	10/02/2022	10/02/2022
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	10/02/2022	10/02/2022
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	10/02/2022	10/02/2022
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	10/02/2022	10/02/2022
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	10/02/2022	10/02/2022
					CT2985	Active	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	10/02/2022	10/02/2022
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	10/02/2022	10/02/2022
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	10/02/2022	10/02/2022
					CT3001	Implantable metal	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	10/02/2022	10/02/2022
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	10/02/2022	10/02/2022
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	10/02/2022	10/02/2022
					CT3002	Implantable synthetic polymer	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	10/02/2022	10/02/2022
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	10/02/2022	10/02/2022

						F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	10/02/2022	10/02/2022
					CT3006	Implantable inorganic material	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	10/02/2022	10/02/2022
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	10/02/2022	10/02/2022
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	10/02/2022	10/02/2022
					CT314	Electrically powered	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	10/02/2022	10/02/2022
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	10/02/2022	10/02/2022
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	10/02/2022	10/02/2022
					CT321	Long-term surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	10/02/2022	10/02/2022
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	10/02/2022	10/02/2022
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	10/02/2022	10/02/2022
					CT449	Generators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	10/02/2022	10/02/2022
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	10/02/2022	10/02/2022
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	10/02/2022	10/02/2022
					CT981	Single use	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	10/02/2022	10/02/2022
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	10/02/2022	10/02/2022
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	10/02/2022	10/02/2022
					CT983	Surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	10/02/2022	10/02/2022
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	10/02/2022	10/02/2022
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	10/02/2022	10/02/2022
10286	2020	All other reportable incidents	37265	Dual-chamber implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E0601	Arrhythmia	A071201	Failure to Convert Rhythm		
							F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A071201	Failure to Convert Rhythm		
							F23	Unexpected Medical Intervention	A071201	Failure to Convert Rhythm		
10287	2020	All other reportable incidents	35852	Single-chamber implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A160601	Premature Elective Replacement Indicator	04/11/2020	04/11/2020
							F1903	Device Explantation	A160601	Premature Elective Replacement Indicator	04/11/2020	04/11/2020
							F1905	Device Revision or Replacement	A160601	Premature Elective Replacement Indicator	04/11/2020	04/11/2020
10288	2020	All other reportable incidents	47265	Dual-chamber implantable pacemaker, rate-responsive	CT141	Electrodes/Leads and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070909	Oversensing	15/07/2020	15/07/2020
							F1905	Device Revision or Replacement	A070909	Oversensing	15/07/2020	15/07/2020
					CT145	Cardiovascular devices	E2403	No Clinical Signs, Symptoms or Conditions	A070909	Oversensing	15/07/2020	15/07/2020
							F1905	Device Revision or Replacement	A070909	Oversensing	15/07/2020	15/07/2020
					CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070909	Oversensing	15/07/2020	15/07/2020
							F1905	Device Revision or Replacement	A070909	Oversensing	15/07/2020	15/07/2020
					CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A070909	Oversensing	15/07/2020	15/07/2020
							F1905	Device Revision or Replacement	A070909	Oversensing	15/07/2020	15/07/2020
					CT233	Surgical	E2403	No Clinical Signs, Symptoms or Conditions	A070909	Oversensing	15/07/2020	15/07/2020
							F1905	Device Revision or Replacement	A070909	Oversensing	15/07/2020	15/07/2020
					CT2406	Implantable	E2403	No Clinical Signs, Symptoms or Conditions	A070909	Oversensing	15/07/2020	15/07/2020
							F1905	Device Revision or Replacement	A070909	Oversensing	15/07/2020	15/07/2020
					CT2407	Active implantable	E2403	No Clinical Signs, Symptoms or Conditions	A070909	Oversensing	15/07/2020	15/07/2020
							F1905	Device Revision or Replacement	A070909	Oversensing	15/07/2020	15/07/2020
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A070909	Oversensing	15/07/2020	15/07/2020
							F1905	Device Revision or Replacement	A070909	Oversensing	15/07/2020	15/07/2020
					CT2727	Active implant devices/ancillaries	E2403	No Clinical Signs, Symptoms or Conditions	A070909	Oversensing	15/07/2020	15/07/2020
							F1905	Device Revision or Replacement	A070909	Oversensing	15/07/2020	15/07/2020
					CT2750	Implantable cardiac electrical stimulators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070909	Oversensing	15/07/2020	15/07/2020
							F1905	Device Revision or Replacement	A070909	Oversensing	15/07/2020	15/07/2020
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A070909	Oversensing	15/07/2020	15/07/2020
							F1905	Device Revision or Replacement	A070909	Oversensing	15/07/2020	15/07/2020
					CT2985	Active	E2403	No Clinical Signs, Symptoms or Conditions	A070909	Oversensing	15/07/2020	15/07/2020
							F1905	Device Revision or Replacement	A070909	Oversensing	15/07/2020	15/07/2020
					CT3001	Implantable metal	E2403	No Clinical Signs, Symptoms or Conditions	A070909	Oversensing	15/07/2020	15/07/2020
							F1905	Device Revision or Replacement	A070909	Oversensing	15/07/2020	15/07/2020
					CT3002	Implantable synthetic polymer	E2403	No Clinical Signs, Symptoms or Conditions	A070909	Oversensing	15/07/2020	15/07/2020
							F1905	Device Revision or Replacement	A070909	Oversensing	15/07/2020	15/07/2020
					CT3006	Implantable inorganic material	E2403	No Clinical Signs, Symptoms or Conditions	A070909	Oversensing	15/07/2020	15/07/2020
							F1905	Device Revision or Replacement	A070909	Oversensing	15/07/2020	15/07/2020
					CT314	Electrically powered	E2403	No Clinical Signs, Symptoms or Conditions	A070909	Oversensing	15/07/2020	15/07/2020
							F1905	Device Revision or Replacement	A070909	Oversensing	15/07/2020	15/07/2020
					CT321	Long-term surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A070909	Oversensing	15/07/2020	15/07/2020
							F1905	Device Revision or Replacement	A070909	Oversensing	15/07/2020	15/07/2020
					CT449	Generators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070909	Oversensing	15/07/2020	15/07/2020
							F1905	Device Revision or Replacement	A070909	Oversensing	15/07/2020	15/07/2020
					CT981	Single use	E2403	No Clinical Signs, Symptoms or Conditions	A070909	Oversensing	15/07/2020	15/07/2020
							F1905	Device Revision or Replacement	A070909	Oversensing	15/07/2020	15/07/2020
					CT983	Surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A070909	Oversensing	15/07/2020	15/07/2020
							F1905	Device Revision or Replacement	A070909	Oversensing	15/07/2020	15/07/2020
10289	2020	All other reportable incidents	36241	Pacemaker/defibrillator lead	CT1684	Cardiac pacemakers and associated devices	E2104	Electric Shock	A071302	Inappropriate/Inadequate Shock/Stimulation		
							E2403	No Clinical Signs, Symptoms or Conditions	A071302	Inappropriate/Inadequate Shock/Stimulation		
							F26	No Health Consequences or Impact	A071302	Inappropriate/Inadequate Shock/Stimulation		
10290	2020	All other reportable incidents	35852	Single-chamber implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E2104	Electric Shock	A071302	Inappropriate/Inadequate Shock/Stimulation		
							F26	No Health Consequences or Impact	A071302	Inappropriate/Inadequate Shock/Stimulation		
10291	2020	All other reportable incidents	35223	Endocardial/interventricular septal pacing lead	CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070909	Oversensing		
									A0722	Impedance Problem		
									A090801	Signal Artifact/Noise		
							F26	No Health Consequences or Impact	A070909	Oversensing		
									A0722	Impedance Problem		
									A090801	Signal Artifact/Noise		
10292	2020	All other reportable incidents	35852	Single-chamber implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	03/12/2020	03/12/2020
							F19	Surgical Intervention	A070504	Premature Discharge of Battery	03/12/2020	03/12/2020
10293	2020	All other reportable incidents	35852	Single-chamber implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	13/11/2020	13/11/2020
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	13/11/2020	13/11/2020
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	13/11/2020	13/11/2020
10294	2020	All other reportable incidents	47267	Single-chamber implantable pacemaker, rate-responsive	CT141	Electrodes/Leads and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT145	Cardiovascular devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		

							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT233	Surgical	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT2406	Implantable	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT2407	Active implantable	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT2727	Active implant devices/ancillaries	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT2750	Implantable cardiac electrical stimulators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT2985	Active	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT3001	Implantable metal	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT3002	Implantable synthetic polymer	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT3006	Implantable inorganic material	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT314	Electrically powered	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT321	Long-term surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT449	Generators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT981	Single use	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT983	Surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
10295	2020	All other reportable incidents	47267	Single-chamber implantable pacemaker, rate-responsive	CT141	Electrodes/Leads and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT145	Cardiovascular devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT233	Surgical	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT2406	Implantable	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT2407	Active implantable	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT2727	Active implant devices/ancillaries	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT2750	Implantable cardiac electrical stimulators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT2985	Active	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT3001	Implantable metal	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT3002	Implantable synthetic polymer	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT3006	Implantable inorganic material	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT314	Electrically powered	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT321	Long-term surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT449	Generators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		

						F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery			
				CT981	Single use	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery			
						F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery			
				CT983	Surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery			
						F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery			
10296	2020	All other reportable incidents	47265	Dual-chamber implantable pacemaker, rate-responsive	CT141	Electrodes/Leads and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A071102	Failure to Interrogate		
							F26	No Health Consequences or Impact	A071102	Failure to Interrogate		
					CT145	Cardiovascular devices	E2403	No Clinical Signs, Symptoms or Conditions	A071102	Failure to Interrogate		
							F26	No Health Consequences or Impact	A071102	Failure to Interrogate		
					CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A071102	Failure to Interrogate		
							F26	No Health Consequences or Impact	A071102	Failure to Interrogate		
					CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A071102	Failure to Interrogate		
							F26	No Health Consequences or Impact	A071102	Failure to Interrogate		
					CT233	Surgical	E2403	No Clinical Signs, Symptoms or Conditions	A071102	Failure to Interrogate		
							F26	No Health Consequences or Impact	A071102	Failure to Interrogate		
					CT2406	Implantable	E2403	No Clinical Signs, Symptoms or Conditions	A071102	Failure to Interrogate		
							F26	No Health Consequences or Impact	A071102	Failure to Interrogate		
					CT2407	Active implantable	E2403	No Clinical Signs, Symptoms or Conditions	A071102	Failure to Interrogate		
							F26	No Health Consequences or Impact	A071102	Failure to Interrogate		
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A071102	Failure to Interrogate		
							F26	No Health Consequences or Impact	A071102	Failure to Interrogate		
					CT2727	Active implant devices/ancillaries	E2403	No Clinical Signs, Symptoms or Conditions	A071102	Failure to Interrogate		
							F26	No Health Consequences or Impact	A071102	Failure to Interrogate		
					CT2750	Implantable cardiac electrical stimulators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A071102	Failure to Interrogate		
							F26	No Health Consequences or Impact	A071102	Failure to Interrogate		
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A071102	Failure to Interrogate		
							F26	No Health Consequences or Impact	A071102	Failure to Interrogate		
					CT2985	Active	E2403	No Clinical Signs, Symptoms or Conditions	A071102	Failure to Interrogate		
							F26	No Health Consequences or Impact	A071102	Failure to Interrogate		
					CT3001	Implantable metal	E2403	No Clinical Signs, Symptoms or Conditions	A071102	Failure to Interrogate		
							F26	No Health Consequences or Impact	A071102	Failure to Interrogate		
					CT3002	Implantable synthetic polymer	E2403	No Clinical Signs, Symptoms or Conditions	A071102	Failure to Interrogate		
							F26	No Health Consequences or Impact	A071102	Failure to Interrogate		
					CT3006	Implantable inorganic material	E2403	No Clinical Signs, Symptoms or Conditions	A071102	Failure to Interrogate		
							F26	No Health Consequences or Impact	A071102	Failure to Interrogate		
					CT314	Electrically powered	E2403	No Clinical Signs, Symptoms or Conditions	A071102	Failure to Interrogate		
							F26	No Health Consequences or Impact	A071102	Failure to Interrogate		
					CT321	Long-term surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A071102	Failure to Interrogate		
							F26	No Health Consequences or Impact	A071102	Failure to Interrogate		
					CT449	Generators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A071102	Failure to Interrogate		
							F26	No Health Consequences or Impact	A071102	Failure to Interrogate		
					CT981	Single use	E2403	No Clinical Signs, Symptoms or Conditions	A071102	Failure to Interrogate		
							F26	No Health Consequences or Impact	A071102	Failure to Interrogate		
					CT983	Surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A071102	Failure to Interrogate		
							F26	No Health Consequences or Impact	A071102	Failure to Interrogate		
10297	2020	All other reportable incidents	35852	Single-chamber implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
10298	2020	All other reportable incidents	35852	Single-chamber implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A072201	High Impedance		
							F26	No Health Consequences or Impact	A072201	High Impedance		
10299	2020	All other reportable incidents	47265	Dual-chamber implantable pacemaker, rate-responsive	CT141	Electrodes/Leads and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	18/01/2021	18/01/2021
							F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A070504	Premature Discharge of Battery	18/01/2021	18/01/2021
							F19	Surgical Intervention	A070504	Premature Discharge of Battery	18/01/2021	18/01/2021
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	18/01/2021	18/01/2021
					CT145	Cardiovascular devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	18/01/2021	18/01/2021
							F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A070504	Premature Discharge of Battery	18/01/2021	18/01/2021
							F19	Surgical Intervention	A070504	Premature Discharge of Battery	18/01/2021	18/01/2021
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	18/01/2021	18/01/2021
					CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	18/01/2021	18/01/2021
							F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A070504	Premature Discharge of Battery	18/01/2021	18/01/2021
							F19	Surgical Intervention	A070504	Premature Discharge of Battery	18/01/2021	18/01/2021
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	18/01/2021	18/01/2021
					CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	18/01/2021	18/01/2021
							F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A070504	Premature Discharge of Battery	18/01/2021	18/01/2021
							F19	Surgical Intervention	A070504	Premature Discharge of Battery	18/01/2021	18/01/2021
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	18/01/2021	18/01/2021
					CT233	Surgical	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	18/01/2021	18/01/2021
							F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A070504	Premature Discharge of Battery	18/01/2021	18/01/2021
							F19	Surgical Intervention	A070504	Premature Discharge of Battery	18/01/2021	18/01/2021
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	18/01/2021	18/01/2021
					CT2406	Implantable	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	18/01/2021	18/01/2021
							F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A070504	Premature Discharge of Battery	18/01/2021	18/01/2021
							F19	Surgical Intervention	A070504	Premature Discharge of Battery	18/01/2021	18/01/2021
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	18/01/2021	18/01/2021
					CT2407	Active implantable	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	18/01/2021	18/01/2021
							F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A070504	Premature Discharge of Battery	18/01/2021	18/01/2021
							F19	Surgical Intervention	A070504	Premature Discharge of Battery	18/01/2021	18/01/2021
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	18/01/2021	18/01/2021
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	18/01/2021	18/01/2021
							F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A070504	Premature Discharge of Battery	18/01/2021	18/01/2021
							F19	Surgical Intervention	A070504	Premature Discharge of Battery	18/01/2021	18/01/2021
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	18/01/2021	18/01/2021
					CT2727	Active implant devices/ancillaries	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	18/01/2021	18/01/2021
							F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A070504	Premature Discharge of Battery	18/01/2021	18/01/2021
							F19	Surgical Intervention	A070504	Premature Discharge of Battery	18/01/2021	18/01/2021
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	18/01/2021	18/01/2021
					CT2750	Implantable cardiac electrical stimulators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	18/01/2021	18/01/2021
							F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A070504	Premature Discharge of Battery	18/01/2021	18/01/2021
							F19	Surgical Intervention	A070504	Premature Discharge of Battery	18/01/2021	18/01/2021

						F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	18/01/2021	18/01/2021
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	18/01/2021	18/01/2021
							F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A070504	Premature Discharge of Battery	18/01/2021	18/01/2021
							F19	Surgical Intervention	A070504	Premature Discharge of Battery	18/01/2021	18/01/2021
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	18/01/2021	18/01/2021
					CT2985	Active	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	18/01/2021	18/01/2021
							F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A070504	Premature Discharge of Battery	18/01/2021	18/01/2021
							F19	Surgical Intervention	A070504	Premature Discharge of Battery	18/01/2021	18/01/2021
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	18/01/2021	18/01/2021
					CT3001	Implantable metal	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	18/01/2021	18/01/2021
							F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A070504	Premature Discharge of Battery	18/01/2021	18/01/2021
							F19	Surgical Intervention	A070504	Premature Discharge of Battery	18/01/2021	18/01/2021
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	18/01/2021	18/01/2021
					CT3002	Implantable synthetic polymer	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	18/01/2021	18/01/2021
							F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A070504	Premature Discharge of Battery	18/01/2021	18/01/2021
							F19	Surgical Intervention	A070504	Premature Discharge of Battery	18/01/2021	18/01/2021
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	18/01/2021	18/01/2021
					CT3006	Implantable inorganic material	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	18/01/2021	18/01/2021
							F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A070504	Premature Discharge of Battery	18/01/2021	18/01/2021
							F19	Surgical Intervention	A070504	Premature Discharge of Battery	18/01/2021	18/01/2021
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	18/01/2021	18/01/2021
					CT314	Electrically powered	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	18/01/2021	18/01/2021
							F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A070504	Premature Discharge of Battery	18/01/2021	18/01/2021
							F19	Surgical Intervention	A070504	Premature Discharge of Battery	18/01/2021	18/01/2021
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	18/01/2021	18/01/2021
					CT321	Long-term surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	18/01/2021	18/01/2021
							F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A070504	Premature Discharge of Battery	18/01/2021	18/01/2021
							F19	Surgical Intervention	A070504	Premature Discharge of Battery	18/01/2021	18/01/2021
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	18/01/2021	18/01/2021
					CT449	Generators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	18/01/2021	18/01/2021
							F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A070504	Premature Discharge of Battery	18/01/2021	18/01/2021
							F19	Surgical Intervention	A070504	Premature Discharge of Battery	18/01/2021	18/01/2021
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	18/01/2021	18/01/2021
					CT981	Single use	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	18/01/2021	18/01/2021
							F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A070504	Premature Discharge of Battery	18/01/2021	18/01/2021
							F19	Surgical Intervention	A070504	Premature Discharge of Battery	18/01/2021	18/01/2021
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	18/01/2021	18/01/2021
					CT983	Surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	18/01/2021	18/01/2021
							F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A070504	Premature Discharge of Battery	18/01/2021	18/01/2021
							F19	Surgical Intervention	A070504	Premature Discharge of Battery	18/01/2021	18/01/2021
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	18/01/2021	18/01/2021
10300	2020	All other reportable incidents	47263	Cardiac resynchronization therapy implantable pacemaker	CT1684	Cardiac pacemakers and associated devices	E0623	Unspecified Heart Problem	A070909	Oversensing		
									A0712	Pacing Problem		
							E2403	No Clinical Signs, Symptoms or Conditions	A070909	Oversensing		
									A0712	Pacing Problem		
							F26	No Health Consequences or Impact	A070909	Oversensing		
									A0712	Pacing Problem		
					CT1685	Cardiac pacemakers	E0623	Unspecified Heart Problem	A070909	Oversensing		
									A0712	Pacing Problem		
							E2403	No Clinical Signs, Symptoms or Conditions	A070909	Oversensing		
									A0712	Pacing Problem		
							F26	No Health Consequences or Impact	A070909	Oversensing		
									A0712	Pacing Problem		
					CT257	Implantable cardiac pacemakers	E0623	Unspecified Heart Problem	A070909	Oversensing		
									A0712	Pacing Problem		
							E2403	No Clinical Signs, Symptoms or Conditions	A070909	Oversensing		
									A0712	Pacing Problem		
							F26	No Health Consequences or Impact	A070909	Oversensing		
									A0712	Pacing Problem		
					CT2866	Implantable cardiac electrical stimulators	E0623	Unspecified Heart Problem	A070909	Oversensing		
									A0712	Pacing Problem		
							E2403	No Clinical Signs, Symptoms or Conditions	A070909	Oversensing		
									A0712	Pacing Problem		
							F26	No Health Consequences or Impact	A070909	Oversensing		
									A0712	Pacing Problem		
10301	2020	All other reportable incidents	36241	Pacemaker/defibrillator lead	CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A040101	Fracture		
									A070101	Failure to Capture		
									A070910	Undersensing		
									A072201	High Impedance		
							F15	Recognized Device or Procedural Complication	A040101	Fracture		
									A070101	Failure to Capture		
									A070910	Undersensing		
									A072201	High Impedance		
10302	2020	All other reportable incidents	47267	Single-chamber implantable pacemaker, rate-responsive	CT141	Electrodes/Leads and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	19/10/2020	19/10/2020
							F19	Surgical Intervention	A070504	Premature Discharge of Battery	19/10/2020	19/10/2020
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	19/10/2020	19/10/2020
					CT145	Cardiovascular devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	19/10/2020	19/10/2020
							F19	Surgical Intervention	A070504	Premature Discharge of Battery	19/10/2020	19/10/2020
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	19/10/2020	19/10/2020
					CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	19/10/2020	19/10/2020
							F19	Surgical Intervention	A070504	Premature Discharge of Battery	19/10/2020	19/10/2020
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	19/10/2020	19/10/2020
					CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	19/10/2020	19/10/2020
							F19	Surgical Intervention	A070504	Premature Discharge of Battery	19/10/2020	19/10/2020
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	19/10/2020	19/10/2020
					CT233	Surgical	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	19/10/2020	19/10/2020
							F19	Surgical Intervention	A070504	Premature Discharge of Battery	19/10/2020	19/10/2020
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	19/10/2020	19/10/2020

					CT2406	Implantable	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	19/10/2020	19/10/2020
							F19	Surgical Intervention	A070504	Premature Discharge of Battery	19/10/2020	19/10/2020
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	19/10/2020	19/10/2020
					CT2407	Active implantable	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	19/10/2020	19/10/2020
							F19	Surgical Intervention	A070504	Premature Discharge of Battery	19/10/2020	19/10/2020
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	19/10/2020	19/10/2020
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	19/10/2020	19/10/2020
							F19	Surgical Intervention	A070504	Premature Discharge of Battery	19/10/2020	19/10/2020
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	19/10/2020	19/10/2020
					CT2727	Active implant devices/ancillaries	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	19/10/2020	19/10/2020
							F19	Surgical Intervention	A070504	Premature Discharge of Battery	19/10/2020	19/10/2020
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	19/10/2020	19/10/2020
					CT2750	Implantable cardiac electrical stimulators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	19/10/2020	19/10/2020
							F19	Surgical Intervention	A070504	Premature Discharge of Battery	19/10/2020	19/10/2020
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	19/10/2020	19/10/2020
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	19/10/2020	19/10/2020
							F19	Surgical Intervention	A070504	Premature Discharge of Battery	19/10/2020	19/10/2020
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	19/10/2020	19/10/2020
					CT2985	Active	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	19/10/2020	19/10/2020
							F19	Surgical Intervention	A070504	Premature Discharge of Battery	19/10/2020	19/10/2020
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	19/10/2020	19/10/2020
					CT3001	Implantable metal	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	19/10/2020	19/10/2020
							F19	Surgical Intervention	A070504	Premature Discharge of Battery	19/10/2020	19/10/2020
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	19/10/2020	19/10/2020
					CT3002	Implantable synthetic polymer	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	19/10/2020	19/10/2020
							F19	Surgical Intervention	A070504	Premature Discharge of Battery	19/10/2020	19/10/2020
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	19/10/2020	19/10/2020
					CT3006	Implantable inorganic material	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	19/10/2020	19/10/2020
							F19	Surgical Intervention	A070504	Premature Discharge of Battery	19/10/2020	19/10/2020
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	19/10/2020	19/10/2020
					CT314	Electrically powered	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	19/10/2020	19/10/2020
							F19	Surgical Intervention	A070504	Premature Discharge of Battery	19/10/2020	19/10/2020
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	19/10/2020	19/10/2020
					CT321	Long-term surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	19/10/2020	19/10/2020
							F19	Surgical Intervention	A070504	Premature Discharge of Battery	19/10/2020	19/10/2020
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	19/10/2020	19/10/2020
					CT449	Generators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	19/10/2020	19/10/2020
							F19	Surgical Intervention	A070504	Premature Discharge of Battery	19/10/2020	19/10/2020
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	19/10/2020	19/10/2020
					CT981	Single use	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	19/10/2020	19/10/2020
							F19	Surgical Intervention	A070504	Premature Discharge of Battery	19/10/2020	19/10/2020
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	19/10/2020	19/10/2020
					CT983	Surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	19/10/2020	19/10/2020
							F19	Surgical Intervention	A070504	Premature Discharge of Battery	19/10/2020	19/10/2020
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	19/10/2020	19/10/2020
10303	2020	All other reportable incidents	47263	Cardiac resynchronization therapy implantable pacemaker	CT1684	Cardiac pacemakers and associated devices	E0506	Hemorrhage/Blood Loss/Bleeding	A24	No Apparent Device Problem		
							E0605	Cardiac Tamponade	A24	No Apparent Device Problem		
							E0619	Pericardial Effusion	A24	No Apparent Device Problem		
							F19	Surgical Intervention	A24	No Apparent Device Problem		
					CT1685	Cardiac pacemakers	E0506	Hemorrhage/Blood Loss/Bleeding	A24	No Apparent Device Problem		
							E0605	Cardiac Tamponade	A24	No Apparent Device Problem		
							E0619	Pericardial Effusion	A24	No Apparent Device Problem		
							F19	Surgical Intervention	A24	No Apparent Device Problem		
					CT257	Implantable cardiac pacemakers	E0506	Hemorrhage/Blood Loss/Bleeding	A24	No Apparent Device Problem		
							E0605	Cardiac Tamponade	A24	No Apparent Device Problem		
							E0619	Pericardial Effusion	A24	No Apparent Device Problem		
							F19	Surgical Intervention	A24	No Apparent Device Problem		
					CT2866	Implantable cardiac electrical stimulators	E0506	Hemorrhage/Blood Loss/Bleeding	A24	No Apparent Device Problem		
							E0605	Cardiac Tamponade	A24	No Apparent Device Problem		
							E0619	Pericardial Effusion	A24	No Apparent Device Problem		
							F19	Surgical Intervention	A24	No Apparent Device Problem		
10304	2020	All other reportable incidents	35852	Single-chamber implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	08/12/2020	08/12/2020
							F1901	Additional Surgery	A070504	Premature Discharge of Battery	08/12/2020	08/12/2020
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	08/12/2020	08/12/2020
10305	2020	All other reportable incidents	47270	Cardiac resynchronization therapy implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	24/11/2020	24/11/2020
							F1901	Additional Surgery	A070504	Premature Discharge of Battery	24/11/2020	24/11/2020
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	24/11/2020	24/11/2020
10306	2020	All other reportable incidents	60789	Intracardiac pacemaker	CT130	Catheters and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A0711	Interrogation Problem		
							F26	No Health Consequences or Impact	A0711	Interrogation Problem		
					CT145	Cardiovascular devices	E2403	No Clinical Signs, Symptoms or Conditions	A0711	Interrogation Problem		
							F26	No Health Consequences or Impact	A0711	Interrogation Problem		
					CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A0711	Interrogation Problem		
							F26	No Health Consequences or Impact	A0711	Interrogation Problem		
					CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A0711	Interrogation Problem		
							F26	No Health Consequences or Impact	A0711	Interrogation Problem		
					CT233	Surgical	E2403	No Clinical Signs, Symptoms or Conditions	A0711	Interrogation Problem		
							F26	No Health Consequences or Impact	A0711	Interrogation Problem		
					CT2406	Implantable	E2403	No Clinical Signs, Symptoms or Conditions	A0711	Interrogation Problem		
							F26	No Health Consequences or Impact	A0711	Interrogation Problem		
					CT2407	Active implantable	E2403	No Clinical Signs, Symptoms or Conditions	A0711	Interrogation Problem		
							F26	No Health Consequences or Impact	A0711	Interrogation Problem		
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A0711	Interrogation Problem		
							F26	No Health Consequences or Impact	A0711	Interrogation Problem		
					CT2727	Active implant devices/ancillaries	E2403	No Clinical Signs, Symptoms or Conditions	A0711	Interrogation Problem		
							F26	No Health Consequences or Impact	A0711	Interrogation Problem		
					CT2750	Implantable cardiac electrical stimulators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A0711	Interrogation Problem		
							F26	No Health Consequences or Impact	A0711	Interrogation Problem		
					CT2818	Catheterization/Cannulation/Fluid path devices	E2403	No Clinical Signs, Symptoms or Conditions	A0711	Interrogation Problem		

						F26	No Health Consequences or Impact	A0711	Interrogation Problem			
				CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A0711	Interrogation Problem			
						F26	No Health Consequences or Impact	A0711	Interrogation Problem			
				CT2985	Active	E2403	No Clinical Signs, Symptoms or Conditions	A0711	Interrogation Problem			
						F26	No Health Consequences or Impact	A0711	Interrogation Problem			
				CT3001	Implantable metal	E2403	No Clinical Signs, Symptoms or Conditions	A0711	Interrogation Problem			
						F26	No Health Consequences or Impact	A0711	Interrogation Problem			
				CT3002	Implantable synthetic polymer	E2403	No Clinical Signs, Symptoms or Conditions	A0711	Interrogation Problem			
						F26	No Health Consequences or Impact	A0711	Interrogation Problem			
				CT3006	Implantable inorganic material	E2403	No Clinical Signs, Symptoms or Conditions	A0711	Interrogation Problem			
						F26	No Health Consequences or Impact	A0711	Interrogation Problem			
				CT314	Electrically powered	E2403	No Clinical Signs, Symptoms or Conditions	A0711	Interrogation Problem			
						F26	No Health Consequences or Impact	A0711	Interrogation Problem			
				CT321	Long-term surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A0711	Interrogation Problem			
						F26	No Health Consequences or Impact	A0711	Interrogation Problem			
				CT981	Single use	E2403	No Clinical Signs, Symptoms or Conditions	A0711	Interrogation Problem			
						F26	No Health Consequences or Impact	A0711	Interrogation Problem			
				CT983	Surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A0711	Interrogation Problem			
						F26	No Health Consequences or Impact	A0711	Interrogation Problem			
10307	2020	All other reportable incidents	35223	Endocardial/interventricular septal pacing lead	CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070909	Oversensing		
									A0712	Pacing Problem		
							F1901	Additional Surgery	A070909	Oversensing		
									A0712	Pacing Problem		
							F1905	Device Revision or Replacement	A070909	Oversensing		
									A0712	Pacing Problem		
10308	2020	All other reportable incidents	37265	Dual-chamber implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	23/11/2020	23/11/2020
							F1901	Additional Surgery	A070504	Premature Discharge of Battery	23/11/2020	23/11/2020
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	23/11/2020	23/11/2020
10309	2020	All other reportable incidents	47265	Dual-chamber implantable pacemaker, rate-responsive	CT141	Electrodes/Leads and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem	10/12/2020	10/12/2020
									A070504	Premature Discharge of Battery	10/12/2020	10/12/2020
							F08	Hospitalization or Prolonged Hospitalization	A0705	Battery Problem	10/12/2020	10/12/2020
									A070504	Premature Discharge of Battery	10/12/2020	10/12/2020
							F1905	Device Revision or Replacement	A0705	Battery Problem	10/12/2020	10/12/2020
									A070504	Premature Discharge of Battery	10/12/2020	10/12/2020
					CT145	Cardiovascular devices	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem	10/12/2020	10/12/2020
									A070504	Premature Discharge of Battery	10/12/2020	10/12/2020
							F08	Hospitalization or Prolonged Hospitalization	A0705	Battery Problem	10/12/2020	10/12/2020
									A070504	Premature Discharge of Battery	10/12/2020	10/12/2020
							F1905	Device Revision or Replacement	A0705	Battery Problem	10/12/2020	10/12/2020
									A070504	Premature Discharge of Battery	10/12/2020	10/12/2020
					CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem	10/12/2020	10/12/2020
									A070504	Premature Discharge of Battery	10/12/2020	10/12/2020
							F08	Hospitalization or Prolonged Hospitalization	A0705	Battery Problem	10/12/2020	10/12/2020
									A070504	Premature Discharge of Battery	10/12/2020	10/12/2020
							F1905	Device Revision or Replacement	A0705	Battery Problem	10/12/2020	10/12/2020
									A070504	Premature Discharge of Battery	10/12/2020	10/12/2020
					CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem	10/12/2020	10/12/2020
									A070504	Premature Discharge of Battery	10/12/2020	10/12/2020
							F08	Hospitalization or Prolonged Hospitalization	A0705	Battery Problem	10/12/2020	10/12/2020
									A070504	Premature Discharge of Battery	10/12/2020	10/12/2020
							F1905	Device Revision or Replacement	A0705	Battery Problem	10/12/2020	10/12/2020
									A070504	Premature Discharge of Battery	10/12/2020	10/12/2020
					CT233	Surgical	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem	10/12/2020	10/12/2020
									A070504	Premature Discharge of Battery	10/12/2020	10/12/2020
							F08	Hospitalization or Prolonged Hospitalization	A0705	Battery Problem	10/12/2020	10/12/2020
									A070504	Premature Discharge of Battery	10/12/2020	10/12/2020
							F1905	Device Revision or Replacement	A0705	Battery Problem	10/12/2020	10/12/2020
									A070504	Premature Discharge of Battery	10/12/2020	10/12/2020
					CT2406	Implantable	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem	10/12/2020	10/12/2020
									A070504	Premature Discharge of Battery	10/12/2020	10/12/2020
							F08	Hospitalization or Prolonged Hospitalization	A0705	Battery Problem	10/12/2020	10/12/2020
									A070504	Premature Discharge of Battery	10/12/2020	10/12/2020
							F1905	Device Revision or Replacement	A0705	Battery Problem	10/12/2020	10/12/2020
									A070504	Premature Discharge of Battery	10/12/2020	10/12/2020
					CT2407	Active implantable	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem	10/12/2020	10/12/2020
									A070504	Premature Discharge of Battery	10/12/2020	10/12/2020
							F08	Hospitalization or Prolonged Hospitalization	A0705	Battery Problem	10/12/2020	10/12/2020
									A070504	Premature Discharge of Battery	10/12/2020	10/12/2020
							F1905	Device Revision or Replacement	A0705	Battery Problem	10/12/2020	10/12/2020
									A070504	Premature Discharge of Battery	10/12/2020	10/12/2020
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem	10/12/2020	10/12/2020
									A070504	Premature Discharge of Battery	10/12/2020	10/12/2020
							F08	Hospitalization or Prolonged Hospitalization	A0705	Battery Problem	10/12/2020	10/12/2020
									A070504	Premature Discharge of Battery	10/12/2020	10/12/2020
							F1905	Device Revision or Replacement	A0705	Battery Problem	10/12/2020	10/12/2020
									A070504	Premature Discharge of Battery	10/12/2020	10/12/2020
					CT2727	Active implant devices/ancillaries	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem	10/12/2020	10/12/2020
									A070504	Premature Discharge of Battery	10/12/2020	10/12/2020
							F08	Hospitalization or Prolonged Hospitalization	A0705	Battery Problem	10/12/2020	10/12/2020
									A070504	Premature Discharge of Battery	10/12/2020	10/12/2020
							F1905	Device Revision or Replacement	A0705	Battery Problem	10/12/2020	10/12/2020
									A070504	Premature Discharge of Battery	10/12/2020	10/12/2020
					CT2750	Implantable cardiac electrical stimulators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem	10/12/2020	10/12/2020
									A070504	Premature Discharge of Battery	10/12/2020	10/12/2020
							F08	Hospitalization or Prolonged Hospitalization	A0705	Battery Problem	10/12/2020	10/12/2020
									A070504	Premature Discharge of Battery	10/12/2020	10/12/2020
							F1905	Device Revision or Replacement	A0705	Battery Problem	10/12/2020	10/12/2020
									A070504	Premature Discharge of Battery	10/12/2020	10/12/2020

					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem	10/12/2020	10/12/2020
									A070504	Premature Discharge of Battery	10/12/2020	10/12/2020
							F08	Hospitalization or Prolonged Hospitalization	A0705	Battery Problem	10/12/2020	10/12/2020
									A070504	Premature Discharge of Battery	10/12/2020	10/12/2020
							F1905	Device Revision or Replacement	A0705	Battery Problem	10/12/2020	10/12/2020
									A070504	Premature Discharge of Battery	10/12/2020	10/12/2020
					CT2985	Active	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem	10/12/2020	10/12/2020
									A070504	Premature Discharge of Battery	10/12/2020	10/12/2020
							F08	Hospitalization or Prolonged Hospitalization	A0705	Battery Problem	10/12/2020	10/12/2020
									A070504	Premature Discharge of Battery	10/12/2020	10/12/2020
							F1905	Device Revision or Replacement	A0705	Battery Problem	10/12/2020	10/12/2020
									A070504	Premature Discharge of Battery	10/12/2020	10/12/2020
					CT3001	Implantable metal	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem	10/12/2020	10/12/2020
									A070504	Premature Discharge of Battery	10/12/2020	10/12/2020
							F08	Hospitalization or Prolonged Hospitalization	A0705	Battery Problem	10/12/2020	10/12/2020
									A070504	Premature Discharge of Battery	10/12/2020	10/12/2020
							F1905	Device Revision or Replacement	A0705	Battery Problem	10/12/2020	10/12/2020
									A070504	Premature Discharge of Battery	10/12/2020	10/12/2020
					CT3002	Implantable synthetic polymer	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem	10/12/2020	10/12/2020
									A070504	Premature Discharge of Battery	10/12/2020	10/12/2020
							F08	Hospitalization or Prolonged Hospitalization	A0705	Battery Problem	10/12/2020	10/12/2020
									A070504	Premature Discharge of Battery	10/12/2020	10/12/2020
							F1905	Device Revision or Replacement	A0705	Battery Problem	10/12/2020	10/12/2020
									A070504	Premature Discharge of Battery	10/12/2020	10/12/2020
					CT3006	Implantable inorganic material	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem	10/12/2020	10/12/2020
									A070504	Premature Discharge of Battery	10/12/2020	10/12/2020
							F08	Hospitalization or Prolonged Hospitalization	A0705	Battery Problem	10/12/2020	10/12/2020
									A070504	Premature Discharge of Battery	10/12/2020	10/12/2020
							F1905	Device Revision or Replacement	A0705	Battery Problem	10/12/2020	10/12/2020
									A070504	Premature Discharge of Battery	10/12/2020	10/12/2020
					CT314	Electrically powered	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem	10/12/2020	10/12/2020
									A070504	Premature Discharge of Battery	10/12/2020	10/12/2020
							F08	Hospitalization or Prolonged Hospitalization	A0705	Battery Problem	10/12/2020	10/12/2020
									A070504	Premature Discharge of Battery	10/12/2020	10/12/2020
							F1905	Device Revision or Replacement	A0705	Battery Problem	10/12/2020	10/12/2020
									A070504	Premature Discharge of Battery	10/12/2020	10/12/2020
					CT321	Long-term surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem	10/12/2020	10/12/2020
									A070504	Premature Discharge of Battery	10/12/2020	10/12/2020
							F08	Hospitalization or Prolonged Hospitalization	A0705	Battery Problem	10/12/2020	10/12/2020
									A070504	Premature Discharge of Battery	10/12/2020	10/12/2020
							F1905	Device Revision or Replacement	A0705	Battery Problem	10/12/2020	10/12/2020
									A070504	Premature Discharge of Battery	10/12/2020	10/12/2020
					CT449	Generators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem	10/12/2020	10/12/2020
									A070504	Premature Discharge of Battery	10/12/2020	10/12/2020
							F08	Hospitalization or Prolonged Hospitalization	A0705	Battery Problem	10/12/2020	10/12/2020
									A070504	Premature Discharge of Battery	10/12/2020	10/12/2020
							F1905	Device Revision or Replacement	A0705	Battery Problem	10/12/2020	10/12/2020
									A070504	Premature Discharge of Battery	10/12/2020	10/12/2020
					CT981	Single use	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem	10/12/2020	10/12/2020
									A070504	Premature Discharge of Battery	10/12/2020	10/12/2020
							F08	Hospitalization or Prolonged Hospitalization	A0705	Battery Problem	10/12/2020	10/12/2020
									A070504	Premature Discharge of Battery	10/12/2020	10/12/2020
							F1905	Device Revision or Replacement	A0705	Battery Problem	10/12/2020	10/12/2020
									A070504	Premature Discharge of Battery	10/12/2020	10/12/2020
					CT983	Surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem	10/12/2020	10/12/2020
									A070504	Premature Discharge of Battery	10/12/2020	10/12/2020
							F08	Hospitalization or Prolonged Hospitalization	A0705	Battery Problem	10/12/2020	10/12/2020
									A070504	Premature Discharge of Battery	10/12/2020	10/12/2020
							F1905	Device Revision or Replacement	A0705	Battery Problem	10/12/2020	10/12/2020
									A070504	Premature Discharge of Battery	10/12/2020	10/12/2020
10310	2020	All other reportable incidents	47265	Dual-chamber implantable pacemaker, rate-responsive	CT141	Electrodes/Leads and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070909	Oversensing		
									A0712	Pacing Problem		
									A072201	High Impedance		
							F15	Recognized Device or Procedural Complication	A070909	Oversensing		
									A0712	Pacing Problem		
									A072201	High Impedance		
					CT145	Cardiovascular devices	E2403	No Clinical Signs, Symptoms or Conditions	A070909	Oversensing		
									A0712	Pacing Problem		
									A072201	High Impedance		
							F15	Recognized Device or Procedural Complication	A070909	Oversensing		
									A0712	Pacing Problem		
									A072201	High Impedance		
					CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070909	Oversensing		
									A0712	Pacing Problem		
									A072201	High Impedance		
							F15	Recognized Device or Procedural Complication	A070909	Oversensing		
									A0712	Pacing Problem		
									A072201	High Impedance		
					CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A070909	Oversensing		
									A0712	Pacing Problem		
									A072201	High Impedance		
							F15	Recognized Device or Procedural Complication	A070909	Oversensing		
									A0712	Pacing Problem		
									A072201	High Impedance		
					CT233	Surgical	E2403	No Clinical Signs, Symptoms or Conditions	A070909	Oversensing		
									A0712	Pacing Problem		
									A072201	High Impedance		
							F15	Recognized Device or Procedural Complication	A070909	Oversensing		

								A0712	Pacing Problem		
								A072201	High Impedance		
					CT2406	Implantable	E2403	No Clinical Signs, Symptoms or Conditions	A070909	Oversensing		
									A0712	Pacing Problem		
									A072201	High Impedance		
							F15	Recognized Device or Procedural Complication	A070909	Oversensing		
									A0712	Pacing Problem		
									A072201	High Impedance		
					CT2407	Active implantable	E2403	No Clinical Signs, Symptoms or Conditions	A070909	Oversensing		
									A0712	Pacing Problem		
									A072201	High Impedance		
							F15	Recognized Device or Procedural Complication	A070909	Oversensing		
									A0712	Pacing Problem		
									A072201	High Impedance		
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A070909	Oversensing		
									A0712	Pacing Problem		
									A072201	High Impedance		
							F15	Recognized Device or Procedural Complication	A070909	Oversensing		
									A0712	Pacing Problem		
									A072201	High Impedance		
					CT2727	Active implant devices/ancillaries	E2403	No Clinical Signs, Symptoms or Conditions	A070909	Oversensing		
									A0712	Pacing Problem		
									A072201	High Impedance		
							F15	Recognized Device or Procedural Complication	A070909	Oversensing		
									A0712	Pacing Problem		
									A072201	High Impedance		
					CT2750	Implantable cardiac electrical stimulators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070909	Oversensing		
									A0712	Pacing Problem		
									A072201	High Impedance		
							F15	Recognized Device or Procedural Complication	A070909	Oversensing		
									A0712	Pacing Problem		
									A072201	High Impedance		
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A070909	Oversensing		
									A0712	Pacing Problem		
									A072201	High Impedance		
							F15	Recognized Device or Procedural Complication	A070909	Oversensing		
									A0712	Pacing Problem		
									A072201	High Impedance		
					CT2985	Active	E2403	No Clinical Signs, Symptoms or Conditions	A070909	Oversensing		
									A0712	Pacing Problem		
									A072201	High Impedance		
							F15	Recognized Device or Procedural Complication	A070909	Oversensing		
									A0712	Pacing Problem		
									A072201	High Impedance		
					CT3001	Implantable metal	E2403	No Clinical Signs, Symptoms or Conditions	A070909	Oversensing		
									A0712	Pacing Problem		
									A072201	High Impedance		
							F15	Recognized Device or Procedural Complication	A070909	Oversensing		
									A0712	Pacing Problem		
									A072201	High Impedance		
					CT3002	Implantable synthetic polymer	E2403	No Clinical Signs, Symptoms or Conditions	A070909	Oversensing		
									A0712	Pacing Problem		
									A072201	High Impedance		
							F15	Recognized Device or Procedural Complication	A070909	Oversensing		
									A0712	Pacing Problem		
									A072201	High Impedance		
					CT3006	Implantable inorganic material	E2403	No Clinical Signs, Symptoms or Conditions	A070909	Oversensing		
									A0712	Pacing Problem		
									A072201	High Impedance		
							F15	Recognized Device or Procedural Complication	A070909	Oversensing		
									A0712	Pacing Problem		
									A072201	High Impedance		
					CT314	Electrically powered	E2403	No Clinical Signs, Symptoms or Conditions	A070909	Oversensing		
									A0712	Pacing Problem		
									A072201	High Impedance		
							F15	Recognized Device or Procedural Complication	A070909	Oversensing		
									A0712	Pacing Problem		
									A072201	High Impedance		
					CT321	Long-term surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A070909	Oversensing		
									A0712	Pacing Problem		
									A072201	High Impedance		
							F15	Recognized Device or Procedural Complication	A070909	Oversensing		
									A0712	Pacing Problem		
									A072201	High Impedance		
					CT449	Generators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070909	Oversensing		
									A0712	Pacing Problem		
									A072201	High Impedance		
							F15	Recognized Device or Procedural Complication	A070909	Oversensing		
									A0712	Pacing Problem		
									A072201	High Impedance		
					CT981	Single use	E2403	No Clinical Signs, Symptoms or Conditions	A070909	Oversensing		
									A0712	Pacing Problem		
									A072201	High Impedance		
							F15	Recognized Device or Procedural Complication	A070909	Oversensing		
									A0712	Pacing Problem		
									A072201	High Impedance		
					CT983	Surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A070909	Oversensing		
									A0712	Pacing Problem		

									A072201	High Impedance		
							F15	Recognized Device or Procedural Complication	A070909	Oversensing		
									A0712	Pacing Problem		
									A072201	High Impedance		
10311	2020	All other reportable incidents	36241	Pacemaker/defibrillator lead	CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070909	Oversensing		
									A0712	Pacing Problem		
									A072201	High Impedance		
							F15	Recognized Device or Procedural Complication	A070909	Oversensing		
									A0712	Pacing Problem		
									A072201	High Impedance		
10312	2020	All other reportable incidents	47265	Dual-chamber implantable pacemaker, rate-responsive	CT141	Electrodes/Leads and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	17/12/2020	17/12/2020
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	17/12/2020	17/12/2020
							F12	Serious Injury/ Illness/ Impairment	A070504	Premature Discharge of Battery	17/12/2020	17/12/2020
							F19	Surgical Intervention	A070504	Premature Discharge of Battery	17/12/2020	17/12/2020
					CT145	Cardiovascular devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	17/12/2020	17/12/2020
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	17/12/2020	17/12/2020
							F12	Serious Injury/ Illness/ Impairment	A070504	Premature Discharge of Battery	17/12/2020	17/12/2020
							F19	Surgical Intervention	A070504	Premature Discharge of Battery	17/12/2020	17/12/2020
					CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	17/12/2020	17/12/2020
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	17/12/2020	17/12/2020
							F12	Serious Injury/ Illness/ Impairment	A070504	Premature Discharge of Battery	17/12/2020	17/12/2020
							F19	Surgical Intervention	A070504	Premature Discharge of Battery	17/12/2020	17/12/2020
					CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	17/12/2020	17/12/2020
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	17/12/2020	17/12/2020
							F12	Serious Injury/ Illness/ Impairment	A070504	Premature Discharge of Battery	17/12/2020	17/12/2020
							F19	Surgical Intervention	A070504	Premature Discharge of Battery	17/12/2020	17/12/2020
					CT233	Surgical	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	17/12/2020	17/12/2020
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	17/12/2020	17/12/2020
							F12	Serious Injury/ Illness/ Impairment	A070504	Premature Discharge of Battery	17/12/2020	17/12/2020
							F19	Surgical Intervention	A070504	Premature Discharge of Battery	17/12/2020	17/12/2020
					CT2406	Implantable	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	17/12/2020	17/12/2020
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	17/12/2020	17/12/2020
							F12	Serious Injury/ Illness/ Impairment	A070504	Premature Discharge of Battery	17/12/2020	17/12/2020
							F19	Surgical Intervention	A070504	Premature Discharge of Battery	17/12/2020	17/12/2020
					CT2407	Active implantable	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	17/12/2020	17/12/2020
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	17/12/2020	17/12/2020
							F12	Serious Injury/ Illness/ Impairment	A070504	Premature Discharge of Battery	17/12/2020	17/12/2020
							F19	Surgical Intervention	A070504	Premature Discharge of Battery	17/12/2020	17/12/2020
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	17/12/2020	17/12/2020
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	17/12/2020	17/12/2020
							F12	Serious Injury/ Illness/ Impairment	A070504	Premature Discharge of Battery	17/12/2020	17/12/2020
							F19	Surgical Intervention	A070504	Premature Discharge of Battery	17/12/2020	17/12/2020
					CT2727	Active implant devices/ancillaries	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	17/12/2020	17/12/2020
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	17/12/2020	17/12/2020
							F12	Serious Injury/ Illness/ Impairment	A070504	Premature Discharge of Battery	17/12/2020	17/12/2020
							F19	Surgical Intervention	A070504	Premature Discharge of Battery	17/12/2020	17/12/2020
					CT2750	Implantable cardiac electrical stimulators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	17/12/2020	17/12/2020
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	17/12/2020	17/12/2020
							F12	Serious Injury/ Illness/ Impairment	A070504	Premature Discharge of Battery	17/12/2020	17/12/2020
							F19	Surgical Intervention	A070504	Premature Discharge of Battery	17/12/2020	17/12/2020
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	17/12/2020	17/12/2020
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	17/12/2020	17/12/2020
							F12	Serious Injury/ Illness/ Impairment	A070504	Premature Discharge of Battery	17/12/2020	17/12/2020
							F19	Surgical Intervention	A070504	Premature Discharge of Battery	17/12/2020	17/12/2020
					CT2985	Active	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	17/12/2020	17/12/2020
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	17/12/2020	17/12/2020
							F12	Serious Injury/ Illness/ Impairment	A070504	Premature Discharge of Battery	17/12/2020	17/12/2020
							F19	Surgical Intervention	A070504	Premature Discharge of Battery	17/12/2020	17/12/2020
					CT3001	Implantable metal	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	17/12/2020	17/12/2020
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	17/12/2020	17/12/2020
							F12	Serious Injury/ Illness/ Impairment	A070504	Premature Discharge of Battery	17/12/2020	17/12/2020
							F19	Surgical Intervention	A070504	Premature Discharge of Battery	17/12/2020	17/12/2020
					CT3002	Implantable synthetic polymer	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	17/12/2020	17/12/2020
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	17/12/2020	17/12/2020
							F12	Serious Injury/ Illness/ Impairment	A070504	Premature Discharge of Battery	17/12/2020	17/12/2020
							F19	Surgical Intervention	A070504	Premature Discharge of Battery	17/12/2020	17/12/2020
					CT3006	Implantable inorganic material	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	17/12/2020	17/12/2020
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	17/12/2020	17/12/2020
							F12	Serious Injury/ Illness/ Impairment	A070504	Premature Discharge of Battery	17/12/2020	17/12/2020
							F19	Surgical Intervention	A070504	Premature Discharge of Battery	17/12/2020	17/12/2020
					CT314	Electrically powered	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	17/12/2020	17/12/2020
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	17/12/2020	17/12/2020
							F12	Serious Injury/ Illness/ Impairment	A070504	Premature Discharge of Battery	17/12/2020	17/12/2020
							F19	Surgical Intervention	A070504	Premature Discharge of Battery	17/12/2020	17/12/2020
					CT321	Long-term surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	17/12/2020	17/12/2020
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	17/12/2020	17/12/2020
							F12	Serious Injury/ Illness/ Impairment	A070504	Premature Discharge of Battery	17/12/2020	17/12/2020
							F19	Surgical Intervention	A070504	Premature Discharge of Battery	17/12/2020	17/12/2020
					CT449	Generators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	17/12/2020	17/12/2020
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	17/12/2020	17/12/2020
							F12	Serious Injury/ Illness/ Impairment	A070504	Premature Discharge of Battery	17/12/2020	17/12/2020
							F19	Surgical Intervention	A070504	Premature Discharge of Battery	17/12/2020	17/12/2020
					CT981	Single use	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	17/12/2020	17/12/2020
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	17/12/2020	17/12/2020
							F12	Serious Injury/ Illness/ Impairment	A070504	Premature Discharge of Battery	17/12/2020	17/12/2020
							F19	Surgical Intervention	A070504	Premature Discharge of Battery	17/12/2020	17/12/2020
					CT983	Surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	17/12/2020	17/12/2020
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	17/12/2020	17/12/2020

							F12	Serious Injury/ Illness/ Impairment	A070504	Premature Discharge of Battery	17/12/2020	17/12/2020
							F19	Surgical Intervention	A070504	Premature Discharge of Battery	17/12/2020	17/12/2020
10313	2020	All other reportable incidents	35852	Single-chamber implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	11/12/2020	11/12/2020
							F1901	Additional Surgery	A070504	Premature Discharge of Battery	11/12/2020	11/12/2020
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	11/12/2020	11/12/2020
10314	2020	All other reportable incidents	60789	Intracardiac pacemaker	CT130	Catheters and associated devices	E0112	Dizziness	A27	Appropriate Device Problem Term/Code not Available		
							E0402	Hypersensitivity/Allergic Reaction	A27	Appropriate Device Problem Term/Code not Available		
							E0601	Arrhythmia	A27	Appropriate Device Problem Term/Code not Available		
							E060104	Bradycardia	A27	Appropriate Device Problem Term/Code not Available		
							E0623	Unspecified Heart Problem	A27	Appropriate Device Problem Term/Code not Available		
							E233001	Chest Pain	A27	Appropriate Device Problem Term/Code not Available		
							F12	Serious Injury/ Illness/ Impairment	A27	Appropriate Device Problem Term/Code not Available		
					CT145	Cardiovascular devices	E0112	Dizziness	A27	Appropriate Device Problem Term/Code not Available		
							E0402	Hypersensitivity/Allergic Reaction	A27	Appropriate Device Problem Term/Code not Available		
							E0601	Arrhythmia	A27	Appropriate Device Problem Term/Code not Available		
							E060104	Bradycardia	A27	Appropriate Device Problem Term/Code not Available		
							E0623	Unspecified Heart Problem	A27	Appropriate Device Problem Term/Code not Available		
							E233001	Chest Pain	A27	Appropriate Device Problem Term/Code not Available		
							F12	Serious Injury/ Illness/ Impairment	A27	Appropriate Device Problem Term/Code not Available		
					CT1684	Cardiac pacemakers and associated devices	E0112	Dizziness	A27	Appropriate Device Problem Term/Code not Available		
							E0402	Hypersensitivity/Allergic Reaction	A27	Appropriate Device Problem Term/Code not Available		
							E0601	Arrhythmia	A27	Appropriate Device Problem Term/Code not Available		
							E060104	Bradycardia	A27	Appropriate Device Problem Term/Code not Available		
							E0623	Unspecified Heart Problem	A27	Appropriate Device Problem Term/Code not Available		
							E233001	Chest Pain	A27	Appropriate Device Problem Term/Code not Available		
							F12	Serious Injury/ Illness/ Impairment	A27	Appropriate Device Problem Term/Code not Available		
					CT1685	Cardiac pacemakers	E0112	Dizziness	A27	Appropriate Device Problem Term/Code not Available		
							E0402	Hypersensitivity/Allergic Reaction	A27	Appropriate Device Problem Term/Code not Available		
							E0601	Arrhythmia	A27	Appropriate Device Problem Term/Code not Available		
							E060104	Bradycardia	A27	Appropriate Device Problem Term/Code not Available		
							E0623	Unspecified Heart Problem	A27	Appropriate Device Problem Term/Code not Available		
							E233001	Chest Pain	A27	Appropriate Device Problem Term/Code not Available		
							F12	Serious Injury/ Illness/ Impairment	A27	Appropriate Device Problem Term/Code not Available		
					CT233	Surgical	E0112	Dizziness	A27	Appropriate Device Problem Term/Code not Available		
							E0402	Hypersensitivity/Allergic Reaction	A27	Appropriate Device Problem Term/Code not Available		
							E0601	Arrhythmia	A27	Appropriate Device Problem Term/Code not Available		
							E060104	Bradycardia	A27	Appropriate Device Problem Term/Code not Available		
							E0623	Unspecified Heart Problem	A27	Appropriate Device Problem Term/Code not Available		
							E233001	Chest Pain	A27	Appropriate Device Problem Term/Code not Available		
							F12	Serious Injury/ Illness/ Impairment	A27	Appropriate Device Problem Term/Code not Available		
					CT2406	Implantable	E0112	Dizziness	A27	Appropriate Device Problem Term/Code not Available		
							E0402	Hypersensitivity/Allergic Reaction	A27	Appropriate Device Problem Term/Code not Available		
							E0601	Arrhythmia	A27	Appropriate Device Problem Term/Code not Available		
							E060104	Bradycardia	A27	Appropriate Device Problem Term/Code not Available		
							E0623	Unspecified Heart Problem	A27	Appropriate Device Problem Term/Code not Available		
							E233001	Chest Pain	A27	Appropriate Device Problem Term/Code not Available		
							F12	Serious Injury/ Illness/ Impairment	A27	Appropriate Device Problem Term/Code not Available		
					CT2407	Active implantable	E0112	Dizziness	A27	Appropriate Device Problem Term/Code not Available		
							E0402	Hypersensitivity/Allergic Reaction	A27	Appropriate Device Problem Term/Code not Available		
							E0601	Arrhythmia	A27	Appropriate Device Problem Term/Code not Available		
							E060104	Bradycardia	A27	Appropriate Device Problem Term/Code not Available		
							E0623	Unspecified Heart Problem	A27	Appropriate Device Problem Term/Code not Available		

							E233001	Chest Pain	A27	Appropriate Device Problem Term/Code not Available		
							F12	Serious Injury/ Illness/ Impairment	A27	Appropriate Device Problem Term/Code not Available		
					CT257	Implantable cardiac pacemakers	E0112	Dizziness	A27	Appropriate Device Problem Term/Code not Available		
							E0402	Hypersensitivity/Allergic Reaction	A27	Appropriate Device Problem Term/Code not Available		
							E0601	Arrhythmia	A27	Appropriate Device Problem Term/Code not Available		
							E060104	Bradycardia	A27	Appropriate Device Problem Term/Code not Available		
							E0623	Unspecified Heart Problem	A27	Appropriate Device Problem Term/Code not Available		
							E233001	Chest Pain	A27	Appropriate Device Problem Term/Code not Available		
							F12	Serious Injury/ Illness/ Impairment	A27	Appropriate Device Problem Term/Code not Available		
					CT2727	Active implant devices/ancillaries	E0112	Dizziness	A27	Appropriate Device Problem Term/Code not Available		
							E0402	Hypersensitivity/Allergic Reaction	A27	Appropriate Device Problem Term/Code not Available		
							E0601	Arrhythmia	A27	Appropriate Device Problem Term/Code not Available		
							E060104	Bradycardia	A27	Appropriate Device Problem Term/Code not Available		
							E0623	Unspecified Heart Problem	A27	Appropriate Device Problem Term/Code not Available		
							E233001	Chest Pain	A27	Appropriate Device Problem Term/Code not Available		
							F12	Serious Injury/ Illness/ Impairment	A27	Appropriate Device Problem Term/Code not Available		
					CT2750	Implantable cardiac electrical stimulators and associated devices	E0112	Dizziness	A27	Appropriate Device Problem Term/Code not Available		
							E0402	Hypersensitivity/Allergic Reaction	A27	Appropriate Device Problem Term/Code not Available		
							E0601	Arrhythmia	A27	Appropriate Device Problem Term/Code not Available		
							E060104	Bradycardia	A27	Appropriate Device Problem Term/Code not Available		
							E0623	Unspecified Heart Problem	A27	Appropriate Device Problem Term/Code not Available		
							E233001	Chest Pain	A27	Appropriate Device Problem Term/Code not Available		
							F12	Serious Injury/ Illness/ Impairment	A27	Appropriate Device Problem Term/Code not Available		
					CT2818	Catheterization/Cannulation/Fluid path devices	E0112	Dizziness	A27	Appropriate Device Problem Term/Code not Available		
							E0402	Hypersensitivity/Allergic Reaction	A27	Appropriate Device Problem Term/Code not Available		
							E0601	Arrhythmia	A27	Appropriate Device Problem Term/Code not Available		
							E060104	Bradycardia	A27	Appropriate Device Problem Term/Code not Available		
							E0623	Unspecified Heart Problem	A27	Appropriate Device Problem Term/Code not Available		
							E233001	Chest Pain	A27	Appropriate Device Problem Term/Code not Available		
							F12	Serious Injury/ Illness/ Impairment	A27	Appropriate Device Problem Term/Code not Available		
					CT2866	Implantable cardiac electrical stimulators	E0112	Dizziness	A27	Appropriate Device Problem Term/Code not Available		
							E0402	Hypersensitivity/Allergic Reaction	A27	Appropriate Device Problem Term/Code not Available		
							E0601	Arrhythmia	A27	Appropriate Device Problem Term/Code not Available		
							E060104	Bradycardia	A27	Appropriate Device Problem Term/Code not Available		
							E0623	Unspecified Heart Problem	A27	Appropriate Device Problem Term/Code not Available		
							E233001	Chest Pain	A27	Appropriate Device Problem Term/Code not Available		
							F12	Serious Injury/ Illness/ Impairment	A27	Appropriate Device Problem Term/Code not Available		
					CT2985	Active	E0112	Dizziness	A27	Appropriate Device Problem Term/Code not Available		
							E0402	Hypersensitivity/Allergic Reaction	A27	Appropriate Device Problem Term/Code not Available		
							E0601	Arrhythmia	A27	Appropriate Device Problem Term/Code not Available		
							E060104	Bradycardia	A27	Appropriate Device Problem Term/Code not Available		
							E0623	Unspecified Heart Problem	A27	Appropriate Device Problem Term/Code not Available		
							E233001	Chest Pain	A27	Appropriate Device Problem Term/Code not Available		
							F12	Serious Injury/ Illness/ Impairment	A27	Appropriate Device Problem Term/Code not Available		
					CT3001	Implantable metal	E0112	Dizziness	A27	Appropriate Device Problem Term/Code not Available		
							E0402	Hypersensitivity/Allergic Reaction	A27	Appropriate Device Problem Term/Code not Available		
							E0601	Arrhythmia	A27	Appropriate Device Problem Term/Code not Available		
							E060104	Bradycardia	A27	Appropriate Device Problem Term/Code not Available		
							E0623	Unspecified Heart Problem	A27	Appropriate Device Problem Term/Code not Available		
							E233001	Chest Pain	A27	Appropriate Device Problem Term/Code not Available		

							F12	Serious Injury/ Illness/ Impairment	A27	Appropriate Device Problem Term/Code not Available		
					CT3002	Implantable synthetic polymer	E0112	Dizziness	A27	Appropriate Device Problem Term/Code not Available		
							E0402	Hypersensitivity/Allergic Reaction	A27	Appropriate Device Problem Term/Code not Available		
							E0601	Arrhythmia	A27	Appropriate Device Problem Term/Code not Available		
							E060104	Bradycardia	A27	Appropriate Device Problem Term/Code not Available		
							E0623	Unspecified Heart Problem	A27	Appropriate Device Problem Term/Code not Available		
							E233001	Chest Pain	A27	Appropriate Device Problem Term/Code not Available		
							F12	Serious Injury/ Illness/ Impairment	A27	Appropriate Device Problem Term/Code not Available		
					CT3006	Implantable inorganic material	E0112	Dizziness	A27	Appropriate Device Problem Term/Code not Available		
							E0402	Hypersensitivity/Allergic Reaction	A27	Appropriate Device Problem Term/Code not Available		
							E0601	Arrhythmia	A27	Appropriate Device Problem Term/Code not Available		
							E060104	Bradycardia	A27	Appropriate Device Problem Term/Code not Available		
							E0623	Unspecified Heart Problem	A27	Appropriate Device Problem Term/Code not Available		
							E233001	Chest Pain	A27	Appropriate Device Problem Term/Code not Available		
							F12	Serious Injury/ Illness/ Impairment	A27	Appropriate Device Problem Term/Code not Available		
					CT314	Electrically powered	E0112	Dizziness	A27	Appropriate Device Problem Term/Code not Available		
							E0402	Hypersensitivity/Allergic Reaction	A27	Appropriate Device Problem Term/Code not Available		
							E0601	Arrhythmia	A27	Appropriate Device Problem Term/Code not Available		
							E060104	Bradycardia	A27	Appropriate Device Problem Term/Code not Available		
							E0623	Unspecified Heart Problem	A27	Appropriate Device Problem Term/Code not Available		
							E233001	Chest Pain	A27	Appropriate Device Problem Term/Code not Available		
							F12	Serious Injury/ Illness/ Impairment	A27	Appropriate Device Problem Term/Code not Available		
					CT321	Long-term surgical invasive	E0112	Dizziness	A27	Appropriate Device Problem Term/Code not Available		
							E0402	Hypersensitivity/Allergic Reaction	A27	Appropriate Device Problem Term/Code not Available		
							E0601	Arrhythmia	A27	Appropriate Device Problem Term/Code not Available		
							E060104	Bradycardia	A27	Appropriate Device Problem Term/Code not Available		
							E0623	Unspecified Heart Problem	A27	Appropriate Device Problem Term/Code not Available		
							E233001	Chest Pain	A27	Appropriate Device Problem Term/Code not Available		
							F12	Serious Injury/ Illness/ Impairment	A27	Appropriate Device Problem Term/Code not Available		
					CT981	Single use	E0112	Dizziness	A27	Appropriate Device Problem Term/Code not Available		
							E0402	Hypersensitivity/Allergic Reaction	A27	Appropriate Device Problem Term/Code not Available		
							E0601	Arrhythmia	A27	Appropriate Device Problem Term/Code not Available		
							E060104	Bradycardia	A27	Appropriate Device Problem Term/Code not Available		
							E0623	Unspecified Heart Problem	A27	Appropriate Device Problem Term/Code not Available		
							E233001	Chest Pain	A27	Appropriate Device Problem Term/Code not Available		
							F12	Serious Injury/ Illness/ Impairment	A27	Appropriate Device Problem Term/Code not Available		
					CT983	Surgical invasive	E0112	Dizziness	A27	Appropriate Device Problem Term/Code not Available		
							E0402	Hypersensitivity/Allergic Reaction	A27	Appropriate Device Problem Term/Code not Available		
							E0601	Arrhythmia	A27	Appropriate Device Problem Term/Code not Available		
							E060104	Bradycardia	A27	Appropriate Device Problem Term/Code not Available		
							E0623	Unspecified Heart Problem	A27	Appropriate Device Problem Term/Code not Available		
							E233001	Chest Pain	A27	Appropriate Device Problem Term/Code not Available		
							F12	Serious Injury/ Illness/ Impairment	A27	Appropriate Device Problem Term/Code not Available		
10315	2020	All other reportable incidents	37265	Dual-chamber implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A0508	Noise, Audible	01/12/2020	01/12/2020
									A070504	Premature Discharge of Battery	01/12/2020	01/12/2020
							F08	Hospitalization or Prolonged Hospitalization	A0508	Noise, Audible	01/12/2020	01/12/2020
									A070504	Premature Discharge of Battery	01/12/2020	01/12/2020
							F12	Serious Injury/ Illness/ Impairment	A0508	Noise, Audible	01/12/2020	01/12/2020
									A070504	Premature Discharge of Battery	01/12/2020	01/12/2020
							F1905	Device Revision or Replacement	A0508	Noise, Audible	01/12/2020	01/12/2020
									A070504	Premature Discharge of Battery	01/12/2020	01/12/2020
10316	2020	All other reportable incidents	36241	Pacemaker/defibrillator lead	CT1684	Cardiac pacemakers and associated devices	E011903	Syncope/Fainting	A051201	Device Dislodged or Dislocated	28/10/2020	28/10/2020
									A070101	Failure to Capture	28/10/2020	28/10/2020
							F19	Surgical Intervention	A051201	Device Dislodged or Dislocated	28/10/2020	28/10/2020
									A070101	Failure to Capture	28/10/2020	28/10/2020

							F1905	Device Revision or Replacement	A051201	Device Dislodged or Dislocated	28/10/2020	28/10/2020
									A070101	Failure to Capture	28/10/2020	28/10/2020
10317	2020	All other reportable incidents	47265	Dual-chamber implantable pacemaker, rate-responsive	CT141	Electrodes/Leads and associated devices	E0601	Arrhythmia	A070101	Failure to Capture		
									A070102	High Capture Threshold		
									A0712	Pacing Problem		
							E0623	Unspecified Heart Problem	A070101	Failure to Capture		
									A070102	High Capture Threshold		
									A0712	Pacing Problem		
							E233001	Chest Pain	A070101	Failure to Capture		
									A070102	High Capture Threshold		
									A0712	Pacing Problem		
							F26	No Health Consequences or Impact	A070101	Failure to Capture		
									A070102	High Capture Threshold		
									A0712	Pacing Problem		
					CT145	Cardiovascular devices	E0601	Arrhythmia	A070101	Failure to Capture		
									A070102	High Capture Threshold		
									A0712	Pacing Problem		
							E0623	Unspecified Heart Problem	A070101	Failure to Capture		
									A070102	High Capture Threshold		
									A0712	Pacing Problem		
							E233001	Chest Pain	A070101	Failure to Capture		
									A070102	High Capture Threshold		
									A0712	Pacing Problem		
							F26	No Health Consequences or Impact	A070101	Failure to Capture		
									A070102	High Capture Threshold		
									A0712	Pacing Problem		
					CT1684	Cardiac pacemakers and associated devices	E0601	Arrhythmia	A070101	Failure to Capture		
									A070102	High Capture Threshold		
									A0712	Pacing Problem		
							E0623	Unspecified Heart Problem	A070101	Failure to Capture		
									A070102	High Capture Threshold		
									A0712	Pacing Problem		
							E233001	Chest Pain	A070101	Failure to Capture		
									A070102	High Capture Threshold		
									A0712	Pacing Problem		
							F26	No Health Consequences or Impact	A070101	Failure to Capture		
									A070102	High Capture Threshold		
									A0712	Pacing Problem		
					CT1685	Cardiac pacemakers	E0601	Arrhythmia	A070101	Failure to Capture		
									A070102	High Capture Threshold		
									A0712	Pacing Problem		
							E0623	Unspecified Heart Problem	A070101	Failure to Capture		
									A070102	High Capture Threshold		
									A0712	Pacing Problem		
							E233001	Chest Pain	A070101	Failure to Capture		
									A070102	High Capture Threshold		
									A0712	Pacing Problem		
							F26	No Health Consequences or Impact	A070101	Failure to Capture		
									A070102	High Capture Threshold		
									A0712	Pacing Problem		
					CT233	Surgical	E0601	Arrhythmia	A070101	Failure to Capture		
									A070102	High Capture Threshold		
									A0712	Pacing Problem		
							E0623	Unspecified Heart Problem	A070101	Failure to Capture		
									A070102	High Capture Threshold		
									A0712	Pacing Problem		
							E233001	Chest Pain	A070101	Failure to Capture		
									A070102	High Capture Threshold		
									A0712	Pacing Problem		
							F26	No Health Consequences or Impact	A070101	Failure to Capture		
									A070102	High Capture Threshold		
									A0712	Pacing Problem		
					CT2406	Implantable	E0601	Arrhythmia	A070101	Failure to Capture		
									A070102	High Capture Threshold		
									A0712	Pacing Problem		
							E0623	Unspecified Heart Problem	A070101	Failure to Capture		
									A070102	High Capture Threshold		
									A0712	Pacing Problem		
							E233001	Chest Pain	A070101	Failure to Capture		
									A070102	High Capture Threshold		
									A0712	Pacing Problem		
							F26	No Health Consequences or Impact	A070101	Failure to Capture		
									A070102	High Capture Threshold		
									A0712	Pacing Problem		
					CT2407	Active implantable	E0601	Arrhythmia	A070101	Failure to Capture		
									A070102	High Capture Threshold		
									A0712	Pacing Problem		
							E0623	Unspecified Heart Problem	A070101	Failure to Capture		
									A070102	High Capture Threshold		
									A0712	Pacing Problem		
							E233001	Chest Pain	A070101	Failure to Capture		
									A070102	High Capture Threshold		
									A0712	Pacing Problem		
							F26	No Health Consequences or Impact	A070101	Failure to Capture		
									A070102	High Capture Threshold		
									A0712	Pacing Problem		
					CT257	Implantable cardiac pacemakers	E0601	Arrhythmia	A070101	Failure to Capture		
									A070102	High Capture Threshold		

									A0712	Pacing Problem		
							E0623	Unspecified Heart Problem	A070101	Failure to Capture		
									A070102	High Capture Threshold		
									A0712	Pacing Problem		
							E233001	Chest Pain	A070101	Failure to Capture		
									A070102	High Capture Threshold		
									A0712	Pacing Problem		
							F26	No Health Consequences or Impact	A070101	Failure to Capture		
									A070102	High Capture Threshold		
									A0712	Pacing Problem		
					CT2727	Active implant devices/ancillaries	E0601	Arrhythmia	A070101	Failure to Capture		
									A070102	High Capture Threshold		
									A0712	Pacing Problem		
							E0623	Unspecified Heart Problem	A070101	Failure to Capture		
									A070102	High Capture Threshold		
									A0712	Pacing Problem		
							E233001	Chest Pain	A070101	Failure to Capture		
									A070102	High Capture Threshold		
									A0712	Pacing Problem		
							F26	No Health Consequences or Impact	A070101	Failure to Capture		
									A070102	High Capture Threshold		
									A0712	Pacing Problem		
					CT2750	Implantable cardiac electrical stimulators and associated devices	E0601	Arrhythmia	A070101	Failure to Capture		
									A070102	High Capture Threshold		
									A0712	Pacing Problem		
							E0623	Unspecified Heart Problem	A070101	Failure to Capture		
									A070102	High Capture Threshold		
									A0712	Pacing Problem		
							E233001	Chest Pain	A070101	Failure to Capture		
									A070102	High Capture Threshold		
									A0712	Pacing Problem		
							F26	No Health Consequences or Impact	A070101	Failure to Capture		
									A070102	High Capture Threshold		
									A0712	Pacing Problem		
					CT2866	Implantable cardiac electrical stimulators	E0601	Arrhythmia	A070101	Failure to Capture		
									A070102	High Capture Threshold		
									A0712	Pacing Problem		
							E0623	Unspecified Heart Problem	A070101	Failure to Capture		
									A070102	High Capture Threshold		
									A0712	Pacing Problem		
							E233001	Chest Pain	A070101	Failure to Capture		
									A070102	High Capture Threshold		
									A0712	Pacing Problem		
							F26	No Health Consequences or Impact	A070101	Failure to Capture		
									A070102	High Capture Threshold		
									A0712	Pacing Problem		
					CT2985	Active	E0601	Arrhythmia	A070101	Failure to Capture		
									A070102	High Capture Threshold		
									A0712	Pacing Problem		
							E0623	Unspecified Heart Problem	A070101	Failure to Capture		
									A070102	High Capture Threshold		
									A0712	Pacing Problem		
							E233001	Chest Pain	A070101	Failure to Capture		
									A070102	High Capture Threshold		
									A0712	Pacing Problem		
							F26	No Health Consequences or Impact	A070101	Failure to Capture		
									A070102	High Capture Threshold		
									A0712	Pacing Problem		
					CT3001	Implantable metal	E0601	Arrhythmia	A070101	Failure to Capture		
									A070102	High Capture Threshold		
									A0712	Pacing Problem		
							E0623	Unspecified Heart Problem	A070101	Failure to Capture		
									A070102	High Capture Threshold		
									A0712	Pacing Problem		
							E233001	Chest Pain	A070101	Failure to Capture		
									A070102	High Capture Threshold		
									A0712	Pacing Problem		
							F26	No Health Consequences or Impact	A070101	Failure to Capture		
									A070102	High Capture Threshold		
									A0712	Pacing Problem		
					CT3002	Implantable synthetic polymer	E0601	Arrhythmia	A070101	Failure to Capture		
									A070102	High Capture Threshold		
									A0712	Pacing Problem		
							E0623	Unspecified Heart Problem	A070101	Failure to Capture		
									A070102	High Capture Threshold		
									A0712	Pacing Problem		
							E233001	Chest Pain	A070101	Failure to Capture		
									A070102	High Capture Threshold		
									A0712	Pacing Problem		
							F26	No Health Consequences or Impact	A070101	Failure to Capture		
									A070102	High Capture Threshold		
									A0712	Pacing Problem		
					CT3006	Implantable inorganic material	E0601	Arrhythmia	A070101	Failure to Capture		
									A070102	High Capture Threshold		
									A0712	Pacing Problem		
							E0623	Unspecified Heart Problem	A070101	Failure to Capture		
									A070102	High Capture Threshold		
									A0712	Pacing Problem		

							E233001	Chest Pain	A070101	Failure to Capture		
									A070102	High Capture Threshold		
									A0712	Pacing Problem		
							F26	No Health Consequences or Impact	A070101	Failure to Capture		
									A070102	High Capture Threshold		
									A0712	Pacing Problem		
					CT314	Electrically powered	E0601	Arrhythmia	A070101	Failure to Capture		
									A070102	High Capture Threshold		
									A0712	Pacing Problem		
							E0623	Unspecified Heart Problem	A070101	Failure to Capture		
									A070102	High Capture Threshold		
									A0712	Pacing Problem		
							E233001	Chest Pain	A070101	Failure to Capture		
									A070102	High Capture Threshold		
									A0712	Pacing Problem		
							F26	No Health Consequences or Impact	A070101	Failure to Capture		
									A070102	High Capture Threshold		
									A0712	Pacing Problem		
					CT321	Long-term surgical invasive	E0601	Arrhythmia	A070101	Failure to Capture		
									A070102	High Capture Threshold		
									A0712	Pacing Problem		
							E0623	Unspecified Heart Problem	A070101	Failure to Capture		
									A070102	High Capture Threshold		
									A0712	Pacing Problem		
							E233001	Chest Pain	A070101	Failure to Capture		
									A070102	High Capture Threshold		
									A0712	Pacing Problem		
							F26	No Health Consequences or Impact	A070101	Failure to Capture		
									A070102	High Capture Threshold		
									A0712	Pacing Problem		
					CT449	Generators and associated devices	E0601	Arrhythmia	A070101	Failure to Capture		
									A070102	High Capture Threshold		
									A0712	Pacing Problem		
							E0623	Unspecified Heart Problem	A070101	Failure to Capture		
									A070102	High Capture Threshold		
									A0712	Pacing Problem		
							E233001	Chest Pain	A070101	Failure to Capture		
									A070102	High Capture Threshold		
									A0712	Pacing Problem		
							F26	No Health Consequences or Impact	A070101	Failure to Capture		
									A070102	High Capture Threshold		
									A0712	Pacing Problem		
					CT981	Single use	E0601	Arrhythmia	A070101	Failure to Capture		
									A070102	High Capture Threshold		
									A0712	Pacing Problem		
							E0623	Unspecified Heart Problem	A070101	Failure to Capture		
									A070102	High Capture Threshold		
									A0712	Pacing Problem		
							E233001	Chest Pain	A070101	Failure to Capture		
									A070102	High Capture Threshold		
									A0712	Pacing Problem		
							F26	No Health Consequences or Impact	A070101	Failure to Capture		
									A070102	High Capture Threshold		
									A0712	Pacing Problem		
					CT983	Surgical invasive	E0601	Arrhythmia	A070101	Failure to Capture		
									A070102	High Capture Threshold		
									A0712	Pacing Problem		
							E0623	Unspecified Heart Problem	A070101	Failure to Capture		
									A070102	High Capture Threshold		
									A0712	Pacing Problem		
							E233001	Chest Pain	A070101	Failure to Capture		
									A070102	High Capture Threshold		
									A0712	Pacing Problem		
							F26	No Health Consequences or Impact	A070101	Failure to Capture		
									A070102	High Capture Threshold		
									A0712	Pacing Problem		
10318	2020	All other reportable incidents	35852	Single-chamber implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
10319	2020	All other reportable incidents	47267	Single-chamber implantable pacemaker, rate-responsive	CT141	Electrodes/Leads and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	02/03/2021	02/03/2021
							F19	Surgical Intervention	A070504	Premature Discharge of Battery	02/03/2021	02/03/2021
					CT145	Cardiovascular devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	02/03/2021	02/03/2021
							F19	Surgical Intervention	A070504	Premature Discharge of Battery	02/03/2021	02/03/2021
					CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	02/03/2021	02/03/2021
							F19	Surgical Intervention	A070504	Premature Discharge of Battery	02/03/2021	02/03/2021
					CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	02/03/2021	02/03/2021
							F19	Surgical Intervention	A070504	Premature Discharge of Battery	02/03/2021	02/03/2021
					CT233	Surgical	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	02/03/2021	02/03/2021
							F19	Surgical Intervention	A070504	Premature Discharge of Battery	02/03/2021	02/03/2021
					CT2406	Implantable	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	02/03/2021	02/03/2021
							F19	Surgical Intervention	A070504	Premature Discharge of Battery	02/03/2021	02/03/2021
					CT2407	Active implantable	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	02/03/2021	02/03/2021
							F19	Surgical Intervention	A070504	Premature Discharge of Battery	02/03/2021	02/03/2021
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	02/03/2021	02/03/2021
							F19	Surgical Intervention	A070504	Premature Discharge of Battery	02/03/2021	02/03/2021
					CT2727	Active implant devices/ancillaries	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	02/03/2021	02/03/2021
							F19	Surgical Intervention	A070504	Premature Discharge of Battery	02/03/2021	02/03/2021
					CT2750	Implantable cardiac electrical stimulators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	02/03/2021	02/03/2021
							F19	Surgical Intervention	A070504	Premature Discharge of Battery	02/03/2021	02/03/2021

					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	02/03/2021	02/03/2021
							F19	Surgical Intervention	A070504	Premature Discharge of Battery	02/03/2021	02/03/2021
					CT2985	Active	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	02/03/2021	02/03/2021
							F19	Surgical Intervention	A070504	Premature Discharge of Battery	02/03/2021	02/03/2021
					CT3001	Implantable metal	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	02/03/2021	02/03/2021
							F19	Surgical Intervention	A070504	Premature Discharge of Battery	02/03/2021	02/03/2021
					CT3002	Implantable synthetic polymer	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	02/03/2021	02/03/2021
							F19	Surgical Intervention	A070504	Premature Discharge of Battery	02/03/2021	02/03/2021
					CT3006	Implantable inorganic material	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	02/03/2021	02/03/2021
							F19	Surgical Intervention	A070504	Premature Discharge of Battery	02/03/2021	02/03/2021
					CT314	Electrically powered	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	02/03/2021	02/03/2021
							F19	Surgical Intervention	A070504	Premature Discharge of Battery	02/03/2021	02/03/2021
					CT321	Long-term surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	02/03/2021	02/03/2021
							F19	Surgical Intervention	A070504	Premature Discharge of Battery	02/03/2021	02/03/2021
					CT449	Generators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	02/03/2021	02/03/2021
							F19	Surgical Intervention	A070504	Premature Discharge of Battery	02/03/2021	02/03/2021
					CT981	Single use	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	02/03/2021	02/03/2021
							F19	Surgical Intervention	A070504	Premature Discharge of Battery	02/03/2021	02/03/2021
					CT983	Surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	02/03/2021	02/03/2021
							F19	Surgical Intervention	A070504	Premature Discharge of Battery	02/03/2021	02/03/2021
10320	2020	All other reportable incidents	35223	Endocardial/interventricular septal pacing lead	CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A0701	Capturing Problem		
									A070102	High Capture Threshold		
									A070901	Decreased Sensitivity		
									A070910	Undersensing		
									A0722	Impedance Problem		
							F26	No Health Consequences or Impact	A0701	Capturing Problem		
									A070102	High Capture Threshold		
									A070901	Decreased Sensitivity		
									A070910	Undersensing		
									A0722	Impedance Problem		
10321	2020	All other reportable incidents	35852	Single-chamber implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E2104	Electric Shock	A071302	Inappropriate/Inadequate Shock/Stimulation		
							F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A071302	Inappropriate/Inadequate Shock/Stimulation		
							F26	No Health Consequences or Impact	A071302	Inappropriate/Inadequate Shock/Stimulation		
10322	2020	All other reportable incidents	36241	Pacemaker/defibrillator lead	CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070102	High Capture Threshold	13/11/2020	13/11/2020
							F19	Surgical Intervention	A070102	High Capture Threshold	13/11/2020	13/11/2020
							F1905	Device Revision or Replacement	A070102	High Capture Threshold	13/11/2020	13/11/2020
10323	2020	All other reportable incidents	47265	Dual-chamber implantable pacemaker, rate-responsive	CT141	Electrodes/Leads and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	23/11/2020	23/11/2020
							F19	Surgical Intervention	A070504	Premature Discharge of Battery	23/11/2020	23/11/2020
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	23/11/2020	23/11/2020
					CT145	Cardiovascular devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	23/11/2020	23/11/2020
							F19	Surgical Intervention	A070504	Premature Discharge of Battery	23/11/2020	23/11/2020
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	23/11/2020	23/11/2020
					CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	23/11/2020	23/11/2020
							F19	Surgical Intervention	A070504	Premature Discharge of Battery	23/11/2020	23/11/2020
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	23/11/2020	23/11/2020
					CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	23/11/2020	23/11/2020
							F19	Surgical Intervention	A070504	Premature Discharge of Battery	23/11/2020	23/11/2020
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	23/11/2020	23/11/2020
					CT233	Surgical	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	23/11/2020	23/11/2020
							F19	Surgical Intervention	A070504	Premature Discharge of Battery	23/11/2020	23/11/2020
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	23/11/2020	23/11/2020
					CT2406	Implantable	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	23/11/2020	23/11/2020
							F19	Surgical Intervention	A070504	Premature Discharge of Battery	23/11/2020	23/11/2020
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	23/11/2020	23/11/2020
					CT2407	Active implantable	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	23/11/2020	23/11/2020
							F19	Surgical Intervention	A070504	Premature Discharge of Battery	23/11/2020	23/11/2020
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	23/11/2020	23/11/2020
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	23/11/2020	23/11/2020
							F19	Surgical Intervention	A070504	Premature Discharge of Battery	23/11/2020	23/11/2020
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	23/11/2020	23/11/2020
					CT2727	Active implant devices/ancillaries	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	23/11/2020	23/11/2020
							F19	Surgical Intervention	A070504	Premature Discharge of Battery	23/11/2020	23/11/2020
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	23/11/2020	23/11/2020
					CT2750	Implantable cardiac electrical stimulators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	23/11/2020	23/11/2020
							F19	Surgical Intervention	A070504	Premature Discharge of Battery	23/11/2020	23/11/2020
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	23/11/2020	23/11/2020
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	23/11/2020	23/11/2020
							F19	Surgical Intervention	A070504	Premature Discharge of Battery	23/11/2020	23/11/2020
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	23/11/2020	23/11/2020
					CT2985	Active	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	23/11/2020	23/11/2020
							F19	Surgical Intervention	A070504	Premature Discharge of Battery	23/11/2020	23/11/2020
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	23/11/2020	23/11/2020
					CT3001	Implantable metal	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	23/11/2020	23/11/2020
							F19	Surgical Intervention	A070504	Premature Discharge of Battery	23/11/2020	23/11/2020
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	23/11/2020	23/11/2020
					CT3002	Implantable synthetic polymer	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	23/11/2020	23/11/2020
							F19	Surgical Intervention	A070504	Premature Discharge of Battery	23/11/2020	23/11/2020
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	23/11/2020	23/11/2020
					CT3006	Implantable inorganic material	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	23/11/2020	23/11/2020
							F19	Surgical Intervention	A070504	Premature Discharge of Battery	23/11/2020	23/11/2020
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	23/11/2020	23/11/2020
					CT314	Electrically powered	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	23/11/2020	23/11/2020
							F19	Surgical Intervention	A070504	Premature Discharge of Battery	23/11/2020	23/11/2020
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	23/11/2020	23/11/2020
					CT321	Long-term surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	23/11/2020	23/11/2020
							F19	Surgical Intervention	A070504	Premature Discharge of Battery	23/11/2020	23/11/2020
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	23/11/2020	23/11/2020
					CT449	Generators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	23/11/2020	23/11/2020

							F19	Surgical Intervention	A070504	Premature Discharge of Battery	23/11/2020	23/11/2020
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	23/11/2020	23/11/2020
					CT981	Single use	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	23/11/2020	23/11/2020
							F19	Surgical Intervention	A070504	Premature Discharge of Battery	23/11/2020	23/11/2020
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	23/11/2020	23/11/2020
					CT983	Surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	23/11/2020	23/11/2020
							F19	Surgical Intervention	A070504	Premature Discharge of Battery	23/11/2020	23/11/2020
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	23/11/2020	23/11/2020
10324	2020	All other reportable incidents	36241	Pacemaker/defibrillator lead	CT1684	Cardiac pacemakers and associated devices	E2114	Perforation	A070101	Failure to Capture		
							F12	Serious Injury/ Illness/ Impairment	A070101	Failure to Capture		
							F1203	Life Threatening Illness or Injury	A070101	Failure to Capture		
							F19	Surgical Intervention	A070101	Failure to Capture		
10325	2020	All other reportable incidents	47265	Dual-chamber implantable pacemaker, rate-responsive	CT141	Electrodes/Leads and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070909	Oversensing		
									A072201	High Impedance		
							F26	No Health Consequences or Impact	A070909	Oversensing		
									A072201	High Impedance		
					CT145	Cardiovascular devices	E2403	No Clinical Signs, Symptoms or Conditions	A070909	Oversensing		
									A072201	High Impedance		
							F26	No Health Consequences or Impact	A070909	Oversensing		
									A072201	High Impedance		
					CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070909	Oversensing		
									A072201	High Impedance		
							F26	No Health Consequences or Impact	A070909	Oversensing		
									A072201	High Impedance		
					CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A070909	Oversensing		
									A072201	High Impedance		
							F26	No Health Consequences or Impact	A070909	Oversensing		
									A072201	High Impedance		
					CT233	Surgical	E2403	No Clinical Signs, Symptoms or Conditions	A070909	Oversensing		
									A072201	High Impedance		
							F26	No Health Consequences or Impact	A070909	Oversensing		
									A072201	High Impedance		
					CT2406	Implantable	E2403	No Clinical Signs, Symptoms or Conditions	A070909	Oversensing		
									A072201	High Impedance		
							F26	No Health Consequences or Impact	A070909	Oversensing		
									A072201	High Impedance		
					CT2407	Active implantable	E2403	No Clinical Signs, Symptoms or Conditions	A070909	Oversensing		
									A072201	High Impedance		
							F26	No Health Consequences or Impact	A070909	Oversensing		
									A072201	High Impedance		
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A070909	Oversensing		
									A072201	High Impedance		
							F26	No Health Consequences or Impact	A070909	Oversensing		
									A072201	High Impedance		
					CT2727	Active implant devices/ancillaries	E2403	No Clinical Signs, Symptoms or Conditions	A070909	Oversensing		
									A072201	High Impedance		
							F26	No Health Consequences or Impact	A070909	Oversensing		
									A072201	High Impedance		
					CT2750	Implantable cardiac electrical stimulators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070909	Oversensing		
									A072201	High Impedance		
							F26	No Health Consequences or Impact	A070909	Oversensing		
									A072201	High Impedance		
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A070909	Oversensing		
									A072201	High Impedance		
							F26	No Health Consequences or Impact	A070909	Oversensing		
									A072201	High Impedance		
					CT2985	Active	E2403	No Clinical Signs, Symptoms or Conditions	A070909	Oversensing		
									A072201	High Impedance		
							F26	No Health Consequences or Impact	A070909	Oversensing		
									A072201	High Impedance		
					CT3001	Implantable metal	E2403	No Clinical Signs, Symptoms or Conditions	A070909	Oversensing		
									A072201	High Impedance		
							F26	No Health Consequences or Impact	A070909	Oversensing		
									A072201	High Impedance		
					CT3002	Implantable synthetic polymer	E2403	No Clinical Signs, Symptoms or Conditions	A070909	Oversensing		
									A072201	High Impedance		
							F26	No Health Consequences or Impact	A070909	Oversensing		
									A072201	High Impedance		
					CT3006	Implantable inorganic material	E2403	No Clinical Signs, Symptoms or Conditions	A070909	Oversensing		
									A072201	High Impedance		
							F26	No Health Consequences or Impact	A070909	Oversensing		
									A072201	High Impedance		
					CT314	Electrically powered	E2403	No Clinical Signs, Symptoms or Conditions	A070909	Oversensing		
									A072201	High Impedance		
							F26	No Health Consequences or Impact	A070909	Oversensing		
									A072201	High Impedance		
					CT321	Long-term surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A070909	Oversensing		
									A072201	High Impedance		
							F26	No Health Consequences or Impact	A070909	Oversensing		
									A072201	High Impedance		
					CT449	Generators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070909	Oversensing		
									A072201	High Impedance		
							F26	No Health Consequences or Impact	A070909	Oversensing		
									A072201	High Impedance		
					CT981	Single use	E2403	No Clinical Signs, Symptoms or Conditions	A070909	Oversensing		
									A072201	High Impedance		
							F26	No Health Consequences or Impact	A070909	Oversensing		
									A072201	High Impedance		

					CT983	Surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A070909	Oversensing		
									A072201	High Impedance		
							F26	No Health Consequences or Impact	A070909	Oversensing		
									A072201	High Impedance		
10326	2020	All other reportable incidents	17882	Physiologic-monitoring defibrillation system	CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070803	Failure to Power Up		
							F27	Problem Identified During Non-Clinical Procedure	A070803	Failure to Power Up		
10327	2020	All other reportable incidents	17882	Physiologic-monitoring defibrillation system	CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A07	Electrical /Electronic Property Problem		
							F27	Problem Identified During Non-Clinical Procedure	A07	Electrical /Electronic Property Problem		
10328	2020	All other reportable incidents	47270	Cardiac resynchronization therapy implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A13	Communication or Transmission Problem		
							F26	No Health Consequences or Impact	A13	Communication or Transmission Problem		
10329	2020	All other reportable incidents	47205	Cardiac pulse generator programmer	CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A13	Communication or Transmission Problem		
							F26	No Health Consequences or Impact	A13	Communication or Transmission Problem		
10330	2020	All other reportable incidents	35852	Single-chamber implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	03/11/2020	03/11/2020
							F19	Surgical Intervention	A070504	Premature Discharge of Battery	03/11/2020	03/11/2020
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	03/11/2020	03/11/2020
10331	2020	All other reportable incidents	47265	Dual-chamber implantable pacemaker, rate-responsive	CT141	Electrodes/Leads and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F24	Insufficient Health Impact Information	A070504	Premature Discharge of Battery		
					CT145	Cardiovascular devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F24	Insufficient Health Impact Information	A070504	Premature Discharge of Battery		
					CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F24	Insufficient Health Impact Information	A070504	Premature Discharge of Battery		
					CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F24	Insufficient Health Impact Information	A070504	Premature Discharge of Battery		
					CT233	Surgical	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F24	Insufficient Health Impact Information	A070504	Premature Discharge of Battery		
					CT2406	Implantable	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F24	Insufficient Health Impact Information	A070504	Premature Discharge of Battery		
					CT2407	Active implantable	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F24	Insufficient Health Impact Information	A070504	Premature Discharge of Battery		
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F24	Insufficient Health Impact Information	A070504	Premature Discharge of Battery		
					CT2727	Active implant devices/ancillaries	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F24	Insufficient Health Impact Information	A070504	Premature Discharge of Battery		
					CT2750	Implantable cardiac electrical stimulators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F24	Insufficient Health Impact Information	A070504	Premature Discharge of Battery		
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F24	Insufficient Health Impact Information	A070504	Premature Discharge of Battery		
					CT2985	Active	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F24	Insufficient Health Impact Information	A070504	Premature Discharge of Battery		
					CT3001	Implantable metal	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F24	Insufficient Health Impact Information	A070504	Premature Discharge of Battery		
					CT3002	Implantable synthetic polymer	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F24	Insufficient Health Impact Information	A070504	Premature Discharge of Battery		
					CT3006	Implantable inorganic material	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F24	Insufficient Health Impact Information	A070504	Premature Discharge of Battery		
					CT314	Electrically powered	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F24	Insufficient Health Impact Information	A070504	Premature Discharge of Battery		
					CT321	Long-term surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F24	Insufficient Health Impact Information	A070504	Premature Discharge of Battery		
					CT449	Generators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F24	Insufficient Health Impact Information	A070504	Premature Discharge of Battery		
					CT981	Single use	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F24	Insufficient Health Impact Information	A070504	Premature Discharge of Battery		
					CT983	Surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F24	Insufficient Health Impact Information	A070504	Premature Discharge of Battery		
10332	2020	All other reportable incidents	35852	Single-chamber implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A070909	Oversensing		
									A072201	High Impedance		
							F26	No Health Consequences or Impact	A070909	Oversensing		
									A072201	High Impedance		
10333	2020	All other reportable incidents	36241	Pacemaker/defibrillator lead	CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070909	Oversensing		
									A072201	High Impedance		
							F26	No Health Consequences or Impact	A070909	Oversensing		
									A072201	High Impedance		
10334	2020	All other reportable incidents	35852	Single-chamber implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A070909	Oversensing		
							F26	No Health Consequences or Impact	A070909	Oversensing		
10335	2020	All other reportable incidents	36241	Pacemaker/defibrillator lead	CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070909	Oversensing		
							F26	No Health Consequences or Impact	A070909	Oversensing		
10336	2020	All other reportable incidents	35852	Single-chamber implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	19/11/2020	19/11/2020
							F19	Surgical Intervention	A070504	Premature Discharge of Battery	19/11/2020	19/11/2020
10337	2020	Death	47265	Dual-chamber implantable pacemaker, rate-responsive	CT141	Electrodes/Leads and associated devices	E2401	Insufficient Information	A24	No Apparent Device Problem		
							F02	Death	A24	No Apparent Device Problem		
					CT145	Cardiovascular devices	E2401	Insufficient Information	A24	No Apparent Device Problem		
							F02	Death	A24	No Apparent Device Problem		
					CT1684	Cardiac pacemakers and associated devices	E2401	Insufficient Information	A24	No Apparent Device Problem		
							F02	Death	A24	No Apparent Device Problem		
					CT1685	Cardiac pacemakers	E2401	Insufficient Information	A24	No Apparent Device Problem		
							F02	Death	A24	No Apparent Device Problem		
					CT233	Surgical	E2401	Insufficient Information	A24	No Apparent Device Problem		
							F02	Death	A24	No Apparent Device Problem		
					CT2406	Implantable	E2401	Insufficient Information	A24	No Apparent Device Problem		
							F02	Death	A24	No Apparent Device Problem		
					CT2407	Active implantable	E2401	Insufficient Information	A24	No Apparent Device Problem		
							F02	Death	A24	No Apparent Device Problem		
					CT257	Implantable cardiac pacemakers	E2401	Insufficient Information	A24	No Apparent Device Problem		
							F02	Death	A24	No Apparent Device Problem		
					CT2727	Active implant devices/ancillaries	E2401	Insufficient Information	A24	No Apparent Device Problem		
							F02	Death	A24	No Apparent Device Problem		
					CT2750	Implantable cardiac electrical stimulators and associated devices	E2401	Insufficient Information	A24	No Apparent Device Problem		

							F02	Death	A24	No Apparent Device Problem		
					CT2866	Implantable cardiac electrical stimulators	E2401	Insufficient Information	A24	No Apparent Device Problem		
							F02	Death	A24	No Apparent Device Problem		
					CT2985	Active	E2401	Insufficient Information	A24	No Apparent Device Problem		
							F02	Death	A24	No Apparent Device Problem		
					CT3001	Implantable metal	E2401	Insufficient Information	A24	No Apparent Device Problem		
							F02	Death	A24	No Apparent Device Problem		
					CT3002	Implantable synthetic polymer	E2401	Insufficient Information	A24	No Apparent Device Problem		
							F02	Death	A24	No Apparent Device Problem		
					CT3006	Implantable inorganic material	E2401	Insufficient Information	A24	No Apparent Device Problem		
							F02	Death	A24	No Apparent Device Problem		
					CT314	Electrically powered	E2401	Insufficient Information	A24	No Apparent Device Problem		
							F02	Death	A24	No Apparent Device Problem		
					CT321	Long-term surgical invasive	E2401	Insufficient Information	A24	No Apparent Device Problem		
							F02	Death	A24	No Apparent Device Problem		
					CT449	Generators and associated devices	E2401	Insufficient Information	A24	No Apparent Device Problem		
							F02	Death	A24	No Apparent Device Problem		
					CT981	Single use	E2401	Insufficient Information	A24	No Apparent Device Problem		
							F02	Death	A24	No Apparent Device Problem		
					CT983	Surgical invasive	E2401	Insufficient Information	A24	No Apparent Device Problem		
							F02	Death	A24	No Apparent Device Problem		
10338	2020	All other reportable incidents	47265	Dual-chamber implantable pacemaker, rate-responsive	CT141	Electrodes/Leads and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070909	Oversensing		
									A0712	Pacing Problem		
									A090801	Signal Artifact/Noise		
							F24	Insufficient Health Impact Information	A070909	Oversensing		
									A0712	Pacing Problem		
									A090801	Signal Artifact/Noise		
					CT145	Cardiovascular devices	E2403	No Clinical Signs, Symptoms or Conditions	A070909	Oversensing		
									A0712	Pacing Problem		
									A090801	Signal Artifact/Noise		
							F24	Insufficient Health Impact Information	A070909	Oversensing		
									A0712	Pacing Problem		
									A090801	Signal Artifact/Noise		
					CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070909	Oversensing		
									A0712	Pacing Problem		
									A090801	Signal Artifact/Noise		
							F24	Insufficient Health Impact Information	A070909	Oversensing		
									A0712	Pacing Problem		
									A090801	Signal Artifact/Noise		
					CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A070909	Oversensing		
									A0712	Pacing Problem		
									A090801	Signal Artifact/Noise		
							F24	Insufficient Health Impact Information	A070909	Oversensing		
									A0712	Pacing Problem		
									A090801	Signal Artifact/Noise		
					CT233	Surgical	E2403	No Clinical Signs, Symptoms or Conditions	A070909	Oversensing		
									A0712	Pacing Problem		
									A090801	Signal Artifact/Noise		
							F24	Insufficient Health Impact Information	A070909	Oversensing		
									A0712	Pacing Problem		
									A090801	Signal Artifact/Noise		
					CT2406	Implantable	E2403	No Clinical Signs, Symptoms or Conditions	A070909	Oversensing		
									A0712	Pacing Problem		
									A090801	Signal Artifact/Noise		
							F24	Insufficient Health Impact Information	A070909	Oversensing		
									A0712	Pacing Problem		
									A090801	Signal Artifact/Noise		
					CT2407	Active implantable	E2403	No Clinical Signs, Symptoms or Conditions	A070909	Oversensing		
									A0712	Pacing Problem		
									A090801	Signal Artifact/Noise		
							F24	Insufficient Health Impact Information	A070909	Oversensing		
									A0712	Pacing Problem		
									A090801	Signal Artifact/Noise		
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A070909	Oversensing		
									A0712	Pacing Problem		
									A090801	Signal Artifact/Noise		
							F24	Insufficient Health Impact Information	A070909	Oversensing		
									A0712	Pacing Problem		
									A090801	Signal Artifact/Noise		
					CT2727	Active implant devices/ancillaries	E2403	No Clinical Signs, Symptoms or Conditions	A070909	Oversensing		
									A0712	Pacing Problem		
									A090801	Signal Artifact/Noise		
							F24	Insufficient Health Impact Information	A070909	Oversensing		
									A0712	Pacing Problem		
									A090801	Signal Artifact/Noise		
					CT2750	Implantable cardiac electrical stimulators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070909	Oversensing		
									A0712	Pacing Problem		
									A090801	Signal Artifact/Noise		
							F24	Insufficient Health Impact Information	A070909	Oversensing		
									A0712	Pacing Problem		
									A090801	Signal Artifact/Noise		
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A070909	Oversensing		
									A0712	Pacing Problem		
									A090801	Signal Artifact/Noise		
							F24	Insufficient Health Impact Information	A070909	Oversensing		
									A0712	Pacing Problem		
									A090801	Signal Artifact/Noise		
					CT2985	Active	E2403	No Clinical Signs, Symptoms or Conditions	A070909	Oversensing		

									A0712	Pacing Problem		
									A090801	Signal Artifact/Noise		
							F24	Insufficient Health Impact Information	A070909	Oversensing		
									A0712	Pacing Problem		
									A090801	Signal Artifact/Noise		
					CT3001	Implantable metal	E2403	No Clinical Signs, Symptoms or Conditions	A070909	Oversensing		
									A0712	Pacing Problem		
									A090801	Signal Artifact/Noise		
							F24	Insufficient Health Impact Information	A070909	Oversensing		
									A0712	Pacing Problem		
									A090801	Signal Artifact/Noise		
					CT3002	Implantable synthetic polymer	E2403	No Clinical Signs, Symptoms or Conditions	A070909	Oversensing		
									A0712	Pacing Problem		
									A090801	Signal Artifact/Noise		
							F24	Insufficient Health Impact Information	A070909	Oversensing		
									A0712	Pacing Problem		
									A090801	Signal Artifact/Noise		
					CT3006	Implantable inorganic material	E2403	No Clinical Signs, Symptoms or Conditions	A070909	Oversensing		
									A0712	Pacing Problem		
									A090801	Signal Artifact/Noise		
							F24	Insufficient Health Impact Information	A070909	Oversensing		
									A0712	Pacing Problem		
									A090801	Signal Artifact/Noise		
					CT314	Electrically powered	E2403	No Clinical Signs, Symptoms or Conditions	A070909	Oversensing		
									A0712	Pacing Problem		
									A090801	Signal Artifact/Noise		
							F24	Insufficient Health Impact Information	A070909	Oversensing		
									A0712	Pacing Problem		
									A090801	Signal Artifact/Noise		
					CT321	Long-term surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A070909	Oversensing		
									A0712	Pacing Problem		
									A090801	Signal Artifact/Noise		
							F24	Insufficient Health Impact Information	A070909	Oversensing		
									A0712	Pacing Problem		
									A090801	Signal Artifact/Noise		
					CT449	Generators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070909	Oversensing		
									A0712	Pacing Problem		
									A090801	Signal Artifact/Noise		
							F24	Insufficient Health Impact Information	A070909	Oversensing		
									A0712	Pacing Problem		
									A090801	Signal Artifact/Noise		
					CT981	Single use	E2403	No Clinical Signs, Symptoms or Conditions	A070909	Oversensing		
									A0712	Pacing Problem		
									A090801	Signal Artifact/Noise		
							F24	Insufficient Health Impact Information	A070909	Oversensing		
									A0712	Pacing Problem		
									A090801	Signal Artifact/Noise		
					CT983	Surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A070909	Oversensing		
									A0712	Pacing Problem		
									A090801	Signal Artifact/Noise		
							F24	Insufficient Health Impact Information	A070909	Oversensing		
									A0712	Pacing Problem		
									A090801	Signal Artifact/Noise		
10339	2020	All other reportable incidents	47267	Single-chamber implantable pacemaker, rate-responsive	CT141	Electrodes/Leads and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	14/01/2021	14/01/2021
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	14/01/2021	14/01/2021
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	14/01/2021	14/01/2021
					CT145	Cardiovascular devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	14/01/2021	14/01/2021
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	14/01/2021	14/01/2021
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	14/01/2021	14/01/2021
					CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	14/01/2021	14/01/2021
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	14/01/2021	14/01/2021
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	14/01/2021	14/01/2021
					CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	14/01/2021	14/01/2021
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	14/01/2021	14/01/2021
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	14/01/2021	14/01/2021
					CT233	Surgical	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	14/01/2021	14/01/2021
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	14/01/2021	14/01/2021
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	14/01/2021	14/01/2021
					CT2406	Implantable	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	14/01/2021	14/01/2021
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	14/01/2021	14/01/2021
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	14/01/2021	14/01/2021
					CT2407	Active implantable	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	14/01/2021	14/01/2021
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	14/01/2021	14/01/2021
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	14/01/2021	14/01/2021
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	14/01/2021	14/01/2021
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	14/01/2021	14/01/2021
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	14/01/2021	14/01/2021
					CT2727	Active implant devices/ancillaries	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	14/01/2021	14/01/2021
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	14/01/2021	14/01/2021
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	14/01/2021	14/01/2021
					CT2750	Implantable cardiac electrical stimulators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	14/01/2021	14/01/2021
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	14/01/2021	14/01/2021
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	14/01/2021	14/01/2021
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	14/01/2021	14/01/2021
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	14/01/2021	14/01/2021
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	14/01/2021	14/01/2021
					CT2985	Active	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	14/01/2021	14/01/2021
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	14/01/2021	14/01/2021

						F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	14/01/2021	14/01/2021
				CT3001	Implantable metal	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	14/01/2021	14/01/2021
						F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	14/01/2021	14/01/2021
						F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	14/01/2021	14/01/2021
				CT3002	Implantable synthetic polymer	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	14/01/2021	14/01/2021
						F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	14/01/2021	14/01/2021
						F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	14/01/2021	14/01/2021
				CT3006	Implantable inorganic material	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	14/01/2021	14/01/2021
						F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	14/01/2021	14/01/2021
						F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	14/01/2021	14/01/2021
				CT314	Electrically powered	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	14/01/2021	14/01/2021
						F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	14/01/2021	14/01/2021
						F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	14/01/2021	14/01/2021
				CT321	Long-term surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	14/01/2021	14/01/2021
						F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	14/01/2021	14/01/2021
						F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	14/01/2021	14/01/2021
				CT449	Generators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	14/01/2021	14/01/2021
						F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	14/01/2021	14/01/2021
						F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	14/01/2021	14/01/2021
				CT981	Single use	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	14/01/2021	14/01/2021
						F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	14/01/2021	14/01/2021
						F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	14/01/2021	14/01/2021
				CT983	Surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	14/01/2021	14/01/2021
						F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	14/01/2021	14/01/2021
						F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	14/01/2021	14/01/2021
10340	2020	All other reportable incidents	47270	Cardiac resynchronization therapy implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem	05/01/2021	05/01/2021
									A070504	Premature Discharge of Battery	05/01/2021	05/01/2021
							F08	Hospitalization or Prolonged Hospitalization	A0705	Battery Problem	05/01/2021	05/01/2021
									A070504	Premature Discharge of Battery	05/01/2021	05/01/2021
							F12	Serious Injury/ Illness/ Impairment	A0705	Battery Problem	05/01/2021	05/01/2021
									A070504	Premature Discharge of Battery	05/01/2021	05/01/2021
							F1905	Device Revision or Replacement	A0705	Battery Problem	05/01/2021	05/01/2021
									A070504	Premature Discharge of Battery	05/01/2021	05/01/2021
10341	2020	All other reportable incidents	47263	Cardiac resynchronization therapy implantable pacemaker	CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	14/11/2020	14/11/2020
							F1901	Additional Surgery	A070504	Premature Discharge of Battery	14/11/2020	14/11/2020
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	14/11/2020	14/11/2020
					CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	14/11/2020	14/11/2020
							F1901	Additional Surgery	A070504	Premature Discharge of Battery	14/11/2020	14/11/2020
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	14/11/2020	14/11/2020
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	14/11/2020	14/11/2020
							F1901	Additional Surgery	A070504	Premature Discharge of Battery	14/11/2020	14/11/2020
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	14/11/2020	14/11/2020
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	14/11/2020	14/11/2020
							F1901	Additional Surgery	A070504	Premature Discharge of Battery	14/11/2020	14/11/2020
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	14/11/2020	14/11/2020
10342	2020	All other reportable incidents	35852	Single-chamber implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E2104	Electric Shock	A071302	Inappropriate/Inadequate Shock/Stimulation		
							E2403	No Clinical Signs, Symptoms or Conditions	A071302	Inappropriate/Inadequate Shock/Stimulation		
							F26	No Health Consequences or Impact	A071302	Inappropriate/Inadequate Shock/Stimulation		
10343	2020	All other reportable incidents	47267	Single-chamber implantable pacemaker, rate-responsive	CT141	Electrodes/Leads and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	16/12/2020	16/12/2020
							F19	Surgical Intervention	A070504	Premature Discharge of Battery	16/12/2020	16/12/2020
					CT145	Cardiovascular devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	16/12/2020	16/12/2020
							F19	Surgical Intervention	A070504	Premature Discharge of Battery	16/12/2020	16/12/2020
					CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	16/12/2020	16/12/2020
							F19	Surgical Intervention	A070504	Premature Discharge of Battery	16/12/2020	16/12/2020
					CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	16/12/2020	16/12/2020
							F19	Surgical Intervention	A070504	Premature Discharge of Battery	16/12/2020	16/12/2020
					CT233	Surgical	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	16/12/2020	16/12/2020
							F19	Surgical Intervention	A070504	Premature Discharge of Battery	16/12/2020	16/12/2020
					CT2406	Implantable	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	16/12/2020	16/12/2020
							F19	Surgical Intervention	A070504	Premature Discharge of Battery	16/12/2020	16/12/2020
					CT2407	Active implantable	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	16/12/2020	16/12/2020
							F19	Surgical Intervention	A070504	Premature Discharge of Battery	16/12/2020	16/12/2020
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	16/12/2020	16/12/2020
							F19	Surgical Intervention	A070504	Premature Discharge of Battery	16/12/2020	16/12/2020
					CT2727	Active implant devices/ancillaries	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	16/12/2020	16/12/2020
							F19	Surgical Intervention	A070504	Premature Discharge of Battery	16/12/2020	16/12/2020
					CT2750	Implantable cardiac electrical stimulators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	16/12/2020	16/12/2020
							F19	Surgical Intervention	A070504	Premature Discharge of Battery	16/12/2020	16/12/2020
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	16/12/2020	16/12/2020
							F19	Surgical Intervention	A070504	Premature Discharge of Battery	16/12/2020	16/12/2020
					CT2985	Active	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	16/12/2020	16/12/2020
							F19	Surgical Intervention	A070504	Premature Discharge of Battery	16/12/2020	16/12/2020
					CT3001	Implantable metal	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	16/12/2020	16/12/2020
							F19	Surgical Intervention	A070504	Premature Discharge of Battery	16/12/2020	16/12/2020
					CT3002	Implantable synthetic polymer	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	16/12/2020	16/12/2020
							F19	Surgical Intervention	A070504	Premature Discharge of Battery	16/12/2020	16/12/2020
					CT3006	Implantable inorganic material	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	16/12/2020	16/12/2020
							F19	Surgical Intervention	A070504	Premature Discharge of Battery	16/12/2020	16/12/2020
					CT314	Electrically powered	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	16/12/2020	16/12/2020
							F19	Surgical Intervention	A070504	Premature Discharge of Battery	16/12/2020	16/12/2020
					CT321	Long-term surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	16/12/2020	16/12/2020
							F19	Surgical Intervention	A070504	Premature Discharge of Battery	16/12/2020	16/12/2020
					CT449	Generators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	16/12/2020	16/12/2020
							F19	Surgical Intervention	A070504	Premature Discharge of Battery	16/12/2020	16/12/2020
					CT981	Single use	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	16/12/2020	16/12/2020
							F19	Surgical Intervention	A070504	Premature Discharge of Battery	16/12/2020	16/12/2020
					CT983	Surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	16/12/2020	16/12/2020
							F19	Surgical Intervention	A070504	Premature Discharge of Battery	16/12/2020	16/12/2020

10344	2020	All other reportable incidents	37265	Dual-chamber implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E060103	Atrial Flutter	A0712	Pacing Problem		
							E2403	No Clinical Signs, Symptoms or Conditions	A0712	Pacing Problem		
							F26	No Health Consequences or Impact	A0712	Pacing Problem		
10345	2020	All other reportable incidents	47265	Dual-chamber implantable pacemaker, rate-responsive	CT141	Electrodes/Leads and associated devices	E2007	Fall	A070909	Oversensing		
									A072202	Low Impedance		
							E2403	No Clinical Signs, Symptoms or Conditions	A070909	Oversensing		
									A072202	Low Impedance		
							F26	No Health Consequences or Impact	A070909	Oversensing		
									A072202	Low Impedance		
					CT145	Cardiovascular devices	E2007	Fall	A070909	Oversensing		
									A072202	Low Impedance		
							E2403	No Clinical Signs, Symptoms or Conditions	A070909	Oversensing		
									A072202	Low Impedance		
							F26	No Health Consequences or Impact	A070909	Oversensing		
									A072202	Low Impedance		
					CT1684	Cardiac pacemakers and associated devices	E2007	Fall	A070909	Oversensing		
									A072202	Low Impedance		
							E2403	No Clinical Signs, Symptoms or Conditions	A070909	Oversensing		
									A072202	Low Impedance		
							F26	No Health Consequences or Impact	A070909	Oversensing		
									A072202	Low Impedance		
					CT1685	Cardiac pacemakers	E2007	Fall	A070909	Oversensing		
									A072202	Low Impedance		
							E2403	No Clinical Signs, Symptoms or Conditions	A070909	Oversensing		
									A072202	Low Impedance		
							F26	No Health Consequences or Impact	A070909	Oversensing		
									A072202	Low Impedance		
					CT233	Surgical	E2007	Fall	A070909	Oversensing		
									A072202	Low Impedance		
							E2403	No Clinical Signs, Symptoms or Conditions	A070909	Oversensing		
									A072202	Low Impedance		
							F26	No Health Consequences or Impact	A070909	Oversensing		
									A072202	Low Impedance		
					CT2406	Implantable	E2007	Fall	A070909	Oversensing		
									A072202	Low Impedance		
							E2403	No Clinical Signs, Symptoms or Conditions	A070909	Oversensing		
									A072202	Low Impedance		
							F26	No Health Consequences or Impact	A070909	Oversensing		
									A072202	Low Impedance		
					CT2407	Active implantable	E2007	Fall	A070909	Oversensing		
									A072202	Low Impedance		
							E2403	No Clinical Signs, Symptoms or Conditions	A070909	Oversensing		
									A072202	Low Impedance		
							F26	No Health Consequences or Impact	A070909	Oversensing		
									A072202	Low Impedance		
					CT257	Implantable cardiac pacemakers	E2007	Fall	A070909	Oversensing		
									A072202	Low Impedance		
							E2403	No Clinical Signs, Symptoms or Conditions	A070909	Oversensing		
									A072202	Low Impedance		
							F26	No Health Consequences or Impact	A070909	Oversensing		
									A072202	Low Impedance		
					CT2727	Active implant devices/ancillaries	E2007	Fall	A070909	Oversensing		
									A072202	Low Impedance		
							E2403	No Clinical Signs, Symptoms or Conditions	A070909	Oversensing		
									A072202	Low Impedance		
							F26	No Health Consequences or Impact	A070909	Oversensing		
									A072202	Low Impedance		
					CT2750	Implantable cardiac electrical stimulators and associated devices	E2007	Fall	A070909	Oversensing		
									A072202	Low Impedance		
							E2403	No Clinical Signs, Symptoms or Conditions	A070909	Oversensing		
									A072202	Low Impedance		
							F26	No Health Consequences or Impact	A070909	Oversensing		
									A072202	Low Impedance		
					CT2866	Implantable cardiac electrical stimulators	E2007	Fall	A070909	Oversensing		
									A072202	Low Impedance		
							E2403	No Clinical Signs, Symptoms or Conditions	A070909	Oversensing		
									A072202	Low Impedance		
							F26	No Health Consequences or Impact	A070909	Oversensing		
									A072202	Low Impedance		
					CT2985	Active	E2007	Fall	A070909	Oversensing		
									A072202	Low Impedance		
							E2403	No Clinical Signs, Symptoms or Conditions	A070909	Oversensing		
									A072202	Low Impedance		
							F26	No Health Consequences or Impact	A070909	Oversensing		
									A072202	Low Impedance		
					CT3001	Implantable metal	E2007	Fall	A070909	Oversensing		
									A072202	Low Impedance		
							E2403	No Clinical Signs, Symptoms or Conditions	A070909	Oversensing		
									A072202	Low Impedance		
							F26	No Health Consequences or Impact	A070909	Oversensing		
									A072202	Low Impedance		
					CT3002	Implantable synthetic polymer	E2007	Fall	A070909	Oversensing		
									A072202	Low Impedance		
							E2403	No Clinical Signs, Symptoms or Conditions	A070909	Oversensing		
									A072202	Low Impedance		
							F26	No Health Consequences or Impact	A070909	Oversensing		
									A072202	Low Impedance		
					CT3006	Implantable inorganic material	E2007	Fall	A070909	Oversensing		

									A072202	Low Impedance		
							E2403	No Clinical Signs, Symptoms or Conditions	A070909	Oversensing		
									A072202	Low Impedance		
							F26	No Health Consequences or Impact	A070909	Oversensing		
									A072202	Low Impedance		
					CT314	Electrically powered	E2007	Fall	A070909	Oversensing		
									A072202	Low Impedance		
							E2403	No Clinical Signs, Symptoms or Conditions	A070909	Oversensing		
									A072202	Low Impedance		
							F26	No Health Consequences or Impact	A070909	Oversensing		
									A072202	Low Impedance		
					CT321	Long-term surgical invasive	E2007	Fall	A070909	Oversensing		
									A072202	Low Impedance		
							E2403	No Clinical Signs, Symptoms or Conditions	A070909	Oversensing		
									A072202	Low Impedance		
							F26	No Health Consequences or Impact	A070909	Oversensing		
									A072202	Low Impedance		
					CT449	Generators and associated devices	E2007	Fall	A070909	Oversensing		
									A072202	Low Impedance		
							E2403	No Clinical Signs, Symptoms or Conditions	A070909	Oversensing		
									A072202	Low Impedance		
							F26	No Health Consequences or Impact	A070909	Oversensing		
									A072202	Low Impedance		
					CT981	Single use	E2007	Fall	A070909	Oversensing		
									A072202	Low Impedance		
							E2403	No Clinical Signs, Symptoms or Conditions	A070909	Oversensing		
									A072202	Low Impedance		
							F26	No Health Consequences or Impact	A070909	Oversensing		
									A072202	Low Impedance		
					CT983	Surgical invasive	E2007	Fall	A070909	Oversensing		
									A072202	Low Impedance		
							E2403	No Clinical Signs, Symptoms or Conditions	A070909	Oversensing		
									A072202	Low Impedance		
							F26	No Health Consequences or Impact	A070909	Oversensing		
									A072202	Low Impedance		
10346	2020	All other reportable incidents	35223	Endocardial/interventricular septal pacing lead	CT1684	Cardiac pacemakers and associated devices	E0518	Unspecified Vascular Problem	A070102	High Capture Threshold	03/08/2012	03/08/2012
									A070909	Oversensing	03/08/2012	03/08/2012
									A072202	Low Impedance	03/08/2012	03/08/2012
									A090801	Signal Artifact/Noise	03/08/2012	03/08/2012
							E2328	Obstruction/Occlusion	A070102	High Capture Threshold	03/08/2012	03/08/2012
									A070909	Oversensing	03/08/2012	03/08/2012
									A072202	Low Impedance	03/08/2012	03/08/2012
									A090801	Signal Artifact/Noise	03/08/2012	03/08/2012
							F1905	Device Revision or Replacement	A070102	High Capture Threshold	03/08/2012	03/08/2012
									A070909	Oversensing	03/08/2012	03/08/2012
									A072202	Low Impedance	03/08/2012	03/08/2012
									A090801	Signal Artifact/Noise	03/08/2012	03/08/2012
10347	2020	All other reportable incidents	35223	Endocardial/interventricular septal pacing lead	CT1684	Cardiac pacemakers and associated devices	E062104	Tricuspid Valve Insufficiency/ Regurgitation	A040101	Fracture		
									A070101	Failure to Capture		
									A070910	Undersensing		
							E0717	Dyspnea	A040101	Fracture		
									A070101	Failure to Capture		
									A070910	Undersensing		
							E0752	Unspecified Respiratory Problem	A040101	Fracture		
									A070101	Failure to Capture		
									A070910	Undersensing		
							F1905	Device Revision or Replacement	A040101	Fracture		
									A070101	Failure to Capture		
									A070910	Undersensing		
10348	2020	All other reportable incidents	17882	Physiologic-monitoring defibrillation system	CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A071301	Failure to Deliver Shock/Stimulation		
							F1001	Absence of Treatment	A071301	Failure to Deliver Shock/Stimulation		
							F26	No Health Consequences or Impact	A071301	Failure to Deliver Shock/Stimulation		
10349	2020	All other reportable incidents	47263	Cardiac resynchronization therapy implantable pacemaker	CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	14/11/2020	14/11/2020
							F1901	Additional Surgery	A070504	Premature Discharge of Battery	14/11/2020	14/11/2020
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	14/11/2020	14/11/2020
					CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	14/11/2020	14/11/2020
							F1901	Additional Surgery	A070504	Premature Discharge of Battery	14/11/2020	14/11/2020
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	14/11/2020	14/11/2020
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	14/11/2020	14/11/2020
							F1901	Additional Surgery	A070504	Premature Discharge of Battery	14/11/2020	14/11/2020
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	14/11/2020	14/11/2020
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	14/11/2020	14/11/2020
							F1901	Additional Surgery	A070504	Premature Discharge of Battery	14/11/2020	14/11/2020
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	14/11/2020	14/11/2020
10350	2020	All other reportable incidents	35852	Single-chamber implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	26/11/2020	26/11/2020
							F1901	Additional Surgery	A070504	Premature Discharge of Battery	26/11/2020	26/11/2020
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	26/11/2020	26/11/2020
10351	2020	All other reportable incidents	37265	Dual-chamber implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
10352	2020	All other reportable incidents	35852	Single-chamber implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A0908	Incorrect, Inadequate or Imprecise Result or Readings	10/11/2020	10/11/2020
							F27	Problem Identified During Non-Clinical Procedure	A0908	Incorrect, Inadequate or Imprecise Result or Readings	10/11/2020	10/11/2020
10353	2020	All other reportable incidents	35852	Single-chamber implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem	22/12/2020	22/12/2020
									A070504	Premature Discharge of Battery	22/12/2020	22/12/2020
							F19	Surgical Intervention	A0705	Battery Problem	22/12/2020	22/12/2020
									A070504	Premature Discharge of Battery	22/12/2020	22/12/2020

10354	2020	All other reportable incidents	47270	Cardiac resynchronization therapy implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	26/11/2020	26/11/2020
							F1901	Additional Surgery	A070504	Premature Discharge of Battery	26/11/2020	26/11/2020
10355	2020	Death	60789	Intracardiac pacemaker	CT130	Catheters and associated devices	E0514	Thrombosis/Thrombus	A070101	Failure to Capture		
									A24	No Apparent Device Problem		
							E0602	Cardiac Arrest	A070101	Failure to Capture		
									A24	No Apparent Device Problem		
							E0604	Cardiac Perforation	A070101	Failure to Capture		
									A24	No Apparent Device Problem		
							E0605	Cardiac Tamponade	A070101	Failure to Capture		
									A24	No Apparent Device Problem		
							E0619	Pericardial Effusion	A070101	Failure to Capture		
									A24	No Apparent Device Problem		
							E2114	Perforation	A070101	Failure to Capture		
									A24	No Apparent Device Problem		
							E2321	Low Blood Pressure/ Hypotension	A070101	Failure to Capture		
									A24	No Apparent Device Problem		
							F02	Death	A070101	Failure to Capture		
									A24	No Apparent Device Problem		
					CT145	Cardiovascular devices	E0514	Thrombosis/Thrombus	A070101	Failure to Capture		
									A24	No Apparent Device Problem		
							E0602	Cardiac Arrest	A070101	Failure to Capture		
									A24	No Apparent Device Problem		
							E0604	Cardiac Perforation	A070101	Failure to Capture		
									A24	No Apparent Device Problem		
							E0605	Cardiac Tamponade	A070101	Failure to Capture		
									A24	No Apparent Device Problem		
							E0619	Pericardial Effusion	A070101	Failure to Capture		
									A24	No Apparent Device Problem		
							E2114	Perforation	A070101	Failure to Capture		
									A24	No Apparent Device Problem		
							E2321	Low Blood Pressure/ Hypotension	A070101	Failure to Capture		
									A24	No Apparent Device Problem		
							F02	Death	A070101	Failure to Capture		
									A24	No Apparent Device Problem		
					CT1684	Cardiac pacemakers and associated devices	E0514	Thrombosis/Thrombus	A070101	Failure to Capture		
									A24	No Apparent Device Problem		
							E0602	Cardiac Arrest	A070101	Failure to Capture		
									A24	No Apparent Device Problem		
							E0604	Cardiac Perforation	A070101	Failure to Capture		
									A24	No Apparent Device Problem		
							E0605	Cardiac Tamponade	A070101	Failure to Capture		
									A24	No Apparent Device Problem		
							E0619	Pericardial Effusion	A070101	Failure to Capture		
									A24	No Apparent Device Problem		
							E2114	Perforation	A070101	Failure to Capture		
									A24	No Apparent Device Problem		
							E2321	Low Blood Pressure/ Hypotension	A070101	Failure to Capture		
									A24	No Apparent Device Problem		
							F02	Death	A070101	Failure to Capture		
									A24	No Apparent Device Problem		
					CT1685	Cardiac pacemakers	E0514	Thrombosis/Thrombus	A070101	Failure to Capture		
									A24	No Apparent Device Problem		
							E0602	Cardiac Arrest	A070101	Failure to Capture		
									A24	No Apparent Device Problem		
							E0604	Cardiac Perforation	A070101	Failure to Capture		
									A24	No Apparent Device Problem		
							E0605	Cardiac Tamponade	A070101	Failure to Capture		
									A24	No Apparent Device Problem		
							E0619	Pericardial Effusion	A070101	Failure to Capture		
									A24	No Apparent Device Problem		
							E2114	Perforation	A070101	Failure to Capture		
									A24	No Apparent Device Problem		
							E2321	Low Blood Pressure/ Hypotension	A070101	Failure to Capture		
									A24	No Apparent Device Problem		
							F02	Death	A070101	Failure to Capture		
									A24	No Apparent Device Problem		
					CT233	Surgical	E0514	Thrombosis/Thrombus	A070101	Failure to Capture		
									A24	No Apparent Device Problem		
							E0602	Cardiac Arrest	A070101	Failure to Capture		
									A24	No Apparent Device Problem		
							E0604	Cardiac Perforation	A070101	Failure to Capture		
									A24	No Apparent Device Problem		
							E0605	Cardiac Tamponade	A070101	Failure to Capture		
									A24	No Apparent Device Problem		
							E0619	Pericardial Effusion	A070101	Failure to Capture		
									A24	No Apparent Device Problem		
							E2114	Perforation	A070101	Failure to Capture		
									A24	No Apparent Device Problem		
							E2321	Low Blood Pressure/ Hypotension	A070101	Failure to Capture		
									A24	No Apparent Device Problem		
							F02	Death	A070101	Failure to Capture		
									A24	No Apparent Device Problem		
					CT2406	Implantable	E0514	Thrombosis/Thrombus	A070101	Failure to Capture		
									A24	No Apparent Device Problem		
							E0602	Cardiac Arrest	A070101	Failure to Capture		
									A24	No Apparent Device Problem		
							E0604	Cardiac Perforation	A070101	Failure to Capture		
									A24	No Apparent Device Problem		

							E0605	Cardiac Tamponade	A070101	Failure to Capture		
									A24	No Apparent Device Problem		
							E0619	Pericardial Effusion	A070101	Failure to Capture		
									A24	No Apparent Device Problem		
							E2114	Perforation	A070101	Failure to Capture		
									A24	No Apparent Device Problem		
							E2321	Low Blood Pressure/ Hypotension	A070101	Failure to Capture		
									A24	No Apparent Device Problem		
							F02	Death	A070101	Failure to Capture		
									A24	No Apparent Device Problem		
					CT2407	Active implantable	E0514	Thrombosis/Thrombus	A070101	Failure to Capture		
									A24	No Apparent Device Problem		
							E0602	Cardiac Arrest	A070101	Failure to Capture		
									A24	No Apparent Device Problem		
							E0604	Cardiac Perforation	A070101	Failure to Capture		
									A24	No Apparent Device Problem		
							E0605	Cardiac Tamponade	A070101	Failure to Capture		
									A24	No Apparent Device Problem		
							E0619	Pericardial Effusion	A070101	Failure to Capture		
									A24	No Apparent Device Problem		
							E2114	Perforation	A070101	Failure to Capture		
									A24	No Apparent Device Problem		
							E2321	Low Blood Pressure/ Hypotension	A070101	Failure to Capture		
									A24	No Apparent Device Problem		
							F02	Death	A070101	Failure to Capture		
									A24	No Apparent Device Problem		
					CT257	Implantable cardiac pacemakers	E0514	Thrombosis/Thrombus	A070101	Failure to Capture		
									A24	No Apparent Device Problem		
							E0602	Cardiac Arrest	A070101	Failure to Capture		
									A24	No Apparent Device Problem		
							E0604	Cardiac Perforation	A070101	Failure to Capture		
									A24	No Apparent Device Problem		
							E0605	Cardiac Tamponade	A070101	Failure to Capture		
									A24	No Apparent Device Problem		
							E0619	Pericardial Effusion	A070101	Failure to Capture		
									A24	No Apparent Device Problem		
							E2114	Perforation	A070101	Failure to Capture		
									A24	No Apparent Device Problem		
							E2321	Low Blood Pressure/ Hypotension	A070101	Failure to Capture		
									A24	No Apparent Device Problem		
							F02	Death	A070101	Failure to Capture		
									A24	No Apparent Device Problem		
					CT2727	Active implant devices/ancillaries	E0514	Thrombosis/Thrombus	A070101	Failure to Capture		
									A24	No Apparent Device Problem		
							E0602	Cardiac Arrest	A070101	Failure to Capture		
									A24	No Apparent Device Problem		
							E0604	Cardiac Perforation	A070101	Failure to Capture		
									A24	No Apparent Device Problem		
							E0605	Cardiac Tamponade	A070101	Failure to Capture		
									A24	No Apparent Device Problem		
							E0619	Pericardial Effusion	A070101	Failure to Capture		
									A24	No Apparent Device Problem		
							E2114	Perforation	A070101	Failure to Capture		
									A24	No Apparent Device Problem		
							E2321	Low Blood Pressure/ Hypotension	A070101	Failure to Capture		
									A24	No Apparent Device Problem		
							F02	Death	A070101	Failure to Capture		
									A24	No Apparent Device Problem		
					CT2750	Implantable cardiac electrical stimulators and associated devices	E0514	Thrombosis/Thrombus	A070101	Failure to Capture		
									A24	No Apparent Device Problem		
							E0602	Cardiac Arrest	A070101	Failure to Capture		
									A24	No Apparent Device Problem		
							E0604	Cardiac Perforation	A070101	Failure to Capture		
									A24	No Apparent Device Problem		
							E0605	Cardiac Tamponade	A070101	Failure to Capture		
									A24	No Apparent Device Problem		
							E0619	Pericardial Effusion	A070101	Failure to Capture		
									A24	No Apparent Device Problem		
							E2114	Perforation	A070101	Failure to Capture		
									A24	No Apparent Device Problem		
							E2321	Low Blood Pressure/ Hypotension	A070101	Failure to Capture		
									A24	No Apparent Device Problem		
							F02	Death	A070101	Failure to Capture		
									A24	No Apparent Device Problem		
					CT2818	Catheterization/Cannulation/Fluid path devices	E0514	Thrombosis/Thrombus	A070101	Failure to Capture		
									A24	No Apparent Device Problem		
							E0602	Cardiac Arrest	A070101	Failure to Capture		
									A24	No Apparent Device Problem		
							E0604	Cardiac Perforation	A070101	Failure to Capture		
									A24	No Apparent Device Problem		
							E0605	Cardiac Tamponade	A070101	Failure to Capture		
									A24	No Apparent Device Problem		
							E0619	Pericardial Effusion	A070101	Failure to Capture		
									A24	No Apparent Device Problem		
							E2114	Perforation	A070101	Failure to Capture		
									A24	No Apparent Device Problem		
							E2321	Low Blood Pressure/ Hypotension	A070101	Failure to Capture		
									A24	No Apparent Device Problem		

							F02	Death	A070101	Failure to Capture		
									A24	No Apparent Device Problem		
					CT2866	Implantable cardiac electrical stimulators	E0514	Thrombosis/Thrombus	A070101	Failure to Capture		
									A24	No Apparent Device Problem		
							E0602	Cardiac Arrest	A070101	Failure to Capture		
									A24	No Apparent Device Problem		
							E0604	Cardiac Perforation	A070101	Failure to Capture		
									A24	No Apparent Device Problem		
							E0605	Cardiac Tamponade	A070101	Failure to Capture		
									A24	No Apparent Device Problem		
							E0619	Pericardial Effusion	A070101	Failure to Capture		
									A24	No Apparent Device Problem		
							E2114	Perforation	A070101	Failure to Capture		
									A24	No Apparent Device Problem		
							E2321	Low Blood Pressure/ Hypotension	A070101	Failure to Capture		
									A24	No Apparent Device Problem		
							F02	Death	A070101	Failure to Capture		
									A24	No Apparent Device Problem		
					CT2985	Active	E0514	Thrombosis/Thrombus	A070101	Failure to Capture		
									A24	No Apparent Device Problem		
							E0602	Cardiac Arrest	A070101	Failure to Capture		
									A24	No Apparent Device Problem		
							E0604	Cardiac Perforation	A070101	Failure to Capture		
									A24	No Apparent Device Problem		
							E0605	Cardiac Tamponade	A070101	Failure to Capture		
									A24	No Apparent Device Problem		
							E0619	Pericardial Effusion	A070101	Failure to Capture		
									A24	No Apparent Device Problem		
							E2114	Perforation	A070101	Failure to Capture		
									A24	No Apparent Device Problem		
							E2321	Low Blood Pressure/ Hypotension	A070101	Failure to Capture		
									A24	No Apparent Device Problem		
							F02	Death	A070101	Failure to Capture		
									A24	No Apparent Device Problem		
					CT3001	Implantable metal	E0514	Thrombosis/Thrombus	A070101	Failure to Capture		
									A24	No Apparent Device Problem		
							E0602	Cardiac Arrest	A070101	Failure to Capture		
									A24	No Apparent Device Problem		
							E0604	Cardiac Perforation	A070101	Failure to Capture		
									A24	No Apparent Device Problem		
							E0605	Cardiac Tamponade	A070101	Failure to Capture		
									A24	No Apparent Device Problem		
							E0619	Pericardial Effusion	A070101	Failure to Capture		
									A24	No Apparent Device Problem		
							E2114	Perforation	A070101	Failure to Capture		
									A24	No Apparent Device Problem		
							E2321	Low Blood Pressure/ Hypotension	A070101	Failure to Capture		
									A24	No Apparent Device Problem		
							F02	Death	A070101	Failure to Capture		
									A24	No Apparent Device Problem		
					CT3002	Implantable synthetic polymer	E0514	Thrombosis/Thrombus	A070101	Failure to Capture		
									A24	No Apparent Device Problem		
							E0602	Cardiac Arrest	A070101	Failure to Capture		
									A24	No Apparent Device Problem		
							E0604	Cardiac Perforation	A070101	Failure to Capture		
									A24	No Apparent Device Problem		
							E0605	Cardiac Tamponade	A070101	Failure to Capture		
									A24	No Apparent Device Problem		
							E0619	Pericardial Effusion	A070101	Failure to Capture		
									A24	No Apparent Device Problem		
							E2114	Perforation	A070101	Failure to Capture		
									A24	No Apparent Device Problem		
							E2321	Low Blood Pressure/ Hypotension	A070101	Failure to Capture		
									A24	No Apparent Device Problem		
							F02	Death	A070101	Failure to Capture		
									A24	No Apparent Device Problem		
					CT3006	Implantable inorganic material	E0514	Thrombosis/Thrombus	A070101	Failure to Capture		
									A24	No Apparent Device Problem		
							E0602	Cardiac Arrest	A070101	Failure to Capture		
									A24	No Apparent Device Problem		
							E0604	Cardiac Perforation	A070101	Failure to Capture		
									A24	No Apparent Device Problem		
							E0605	Cardiac Tamponade	A070101	Failure to Capture		
									A24	No Apparent Device Problem		
							E0619	Pericardial Effusion	A070101	Failure to Capture		
									A24	No Apparent Device Problem		
							E2114	Perforation	A070101	Failure to Capture		
									A24	No Apparent Device Problem		
							E2321	Low Blood Pressure/ Hypotension	A070101	Failure to Capture		
									A24	No Apparent Device Problem		
							F02	Death	A070101	Failure to Capture		
									A24	No Apparent Device Problem		
					CT314	Electrically powered	E0514	Thrombosis/Thrombus	A070101	Failure to Capture		
									A24	No Apparent Device Problem		
							E0602	Cardiac Arrest	A070101	Failure to Capture		
									A24	No Apparent Device Problem		
							E0604	Cardiac Perforation	A070101	Failure to Capture		
									A24	No Apparent Device Problem		

							E0605	Cardiac Tamponade	A070101	Failure to Capture		
									A24	No Apparent Device Problem		
							E0619	Pericardial Effusion	A070101	Failure to Capture		
									A24	No Apparent Device Problem		
							E2114	Perforation	A070101	Failure to Capture		
									A24	No Apparent Device Problem		
							E2321	Low Blood Pressure/ Hypotension	A070101	Failure to Capture		
									A24	No Apparent Device Problem		
							F02	Death	A070101	Failure to Capture		
									A24	No Apparent Device Problem		
					CT321	Long-term surgical invasive	E0514	Thrombosis/Thrombus	A070101	Failure to Capture		
									A24	No Apparent Device Problem		
							E0602	Cardiac Arrest	A070101	Failure to Capture		
									A24	No Apparent Device Problem		
							E0604	Cardiac Perforation	A070101	Failure to Capture		
									A24	No Apparent Device Problem		
							E0605	Cardiac Tamponade	A070101	Failure to Capture		
									A24	No Apparent Device Problem		
							E0619	Pericardial Effusion	A070101	Failure to Capture		
									A24	No Apparent Device Problem		
							E2114	Perforation	A070101	Failure to Capture		
									A24	No Apparent Device Problem		
							E2321	Low Blood Pressure/ Hypotension	A070101	Failure to Capture		
									A24	No Apparent Device Problem		
							F02	Death	A070101	Failure to Capture		
									A24	No Apparent Device Problem		
					CT981	Single use	E0514	Thrombosis/Thrombus	A070101	Failure to Capture		
									A24	No Apparent Device Problem		
							E0602	Cardiac Arrest	A070101	Failure to Capture		
									A24	No Apparent Device Problem		
							E0604	Cardiac Perforation	A070101	Failure to Capture		
									A24	No Apparent Device Problem		
							E0605	Cardiac Tamponade	A070101	Failure to Capture		
									A24	No Apparent Device Problem		
							E0619	Pericardial Effusion	A070101	Failure to Capture		
									A24	No Apparent Device Problem		
							E2114	Perforation	A070101	Failure to Capture		
									A24	No Apparent Device Problem		
							E2321	Low Blood Pressure/ Hypotension	A070101	Failure to Capture		
									A24	No Apparent Device Problem		
							F02	Death	A070101	Failure to Capture		
									A24	No Apparent Device Problem		
					CT983	Surgical invasive	E0514	Thrombosis/Thrombus	A070101	Failure to Capture		
									A24	No Apparent Device Problem		
							E0602	Cardiac Arrest	A070101	Failure to Capture		
									A24	No Apparent Device Problem		
							E0604	Cardiac Perforation	A070101	Failure to Capture		
									A24	No Apparent Device Problem		
							E0605	Cardiac Tamponade	A070101	Failure to Capture		
									A24	No Apparent Device Problem		
							E0619	Pericardial Effusion	A070101	Failure to Capture		
									A24	No Apparent Device Problem		
							E2114	Perforation	A070101	Failure to Capture		
									A24	No Apparent Device Problem		
							E2321	Low Blood Pressure/ Hypotension	A070101	Failure to Capture		
									A24	No Apparent Device Problem		
							F02	Death	A070101	Failure to Capture		
									A24	No Apparent Device Problem		
10356	2020	All other reportable incidents	35852	Single-chamber implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A072201	High Impedance		
							F26	No Health Consequences or Impact	A072201	High Impedance		
10357	2020	All other reportable incidents	37265	Dual-chamber implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A070909	Oversensing		
									A072201	High Impedance		
							F26	No Health Consequences or Impact	A070909	Oversensing		
									A072201	High Impedance		
10358	2020	All other reportable incidents	47265	Dual-chamber implantable pacemaker, rate-responsive	CT141	Electrodes/Leads and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT145	Cardiovascular devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT233	Surgical	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT2406	Implantable	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT2407	Active implantable	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT2727	Active implant devices/ancillaries	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT2750	Implantable cardiac electrical stimulators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT2985	Active	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		

					CT3001	Implantable metal	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT3002	Implantable synthetic polymer	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT3006	Implantable inorganic material	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT314	Electrically powered	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT321	Long-term surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT449	Generators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT981	Single use	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT983	Surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
10359	2020	All other reportable incidents	47270	Cardiac resynchronization therapy implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E2104	Electric Shock	A071302	Inappropriate/Inadequate Shock/Stimulation		
							F26	No Health Consequences or Impact	A071302	Inappropriate/Inadequate Shock/Stimulation		
10360	2020	All other reportable incidents	36241	Pacemaker/defibrillator lead	CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A072201	High Impedance		
							F26	No Health Consequences or Impact	A072201	High Impedance		
10361	2020	Death	37265	Dual-chamber implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E0601	Arrhythmia	A0908	Incorrect, Inadequate or Imprecise Result or Readings	29/11/2020	29/11/2020
							F02	Death	A0908	Incorrect, Inadequate or Imprecise Result or Readings	29/11/2020	29/11/2020
10362	2020	All other reportable incidents	35852	Single-chamber implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E2104	Electric Shock	A070909	Oversensing		
									A071302	Inappropriate/Inadequate Shock/Stimulation		
							F08	Hospitalization or Prolonged Hospitalization	A070909	Oversensing		
									A071302	Inappropriate/Inadequate Shock/Stimulation		
							F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A070909	Oversensing		
									A071302	Inappropriate/Inadequate Shock/Stimulation		
							F15	Recognized Device or Procedural Complication	A070909	Oversensing		
									A071302	Inappropriate/Inadequate Shock/Stimulation		
10363	2020	All other reportable incidents	47267	Single-chamber implantable pacemaker, rate-responsive	CT141	Electrodes/Leads and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT145	Cardiovascular devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT233	Surgical	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT2406	Implantable	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT2407	Active implantable	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT2727	Active implant devices/ancillaries	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT2750	Implantable cardiac electrical stimulators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT2985	Active	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT3001	Implantable metal	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT3002	Implantable synthetic polymer	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT3006	Implantable inorganic material	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT314	Electrically powered	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT321	Long-term surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT449	Generators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT981	Single use	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery		

							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT983	Surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
10364	2020	All other reportable incidents	35852	Single-chamber implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A070909	Oversensing		
							F26	No Health Consequences or Impact	A070909	Oversensing		
10365	2020	All other reportable incidents	37265	Dual-chamber implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A070909	Oversensing		
							F26	No Health Consequences or Impact	A070909	Oversensing		
10366	2020	All other reportable incidents	36241	Pacemaker/defibrillator lead	CT1684	Cardiac pacemakers and associated devices	E2104	Electric Shock	A0203	Defective Device	16/04/2021	16/04/2021
									A070909	Oversensing	16/04/2021	16/04/2021
									A071302	Inappropriate/Inadequate Shock/Stimulation	16/04/2021	16/04/2021
							F08	Hospitalization or Prolonged Hospitalization	A0203	Defective Device	16/04/2021	16/04/2021
									A070909	Oversensing	16/04/2021	16/04/2021
									A071302	Inappropriate/Inadequate Shock/Stimulation	16/04/2021	16/04/2021
							F12	Serious Injury/ Illness/ Impairment	A0203	Defective Device	16/04/2021	16/04/2021
									A070909	Oversensing	16/04/2021	16/04/2021
									A071302	Inappropriate/Inadequate Shock/Stimulation	16/04/2021	16/04/2021
							F14	Prolonged Episode of Care	A0203	Defective Device	16/04/2021	16/04/2021
									A070909	Oversensing	16/04/2021	16/04/2021
									A071302	Inappropriate/Inadequate Shock/Stimulation	16/04/2021	16/04/2021
							F1905	Device Revision or Replacement	A0203	Defective Device	16/04/2021	16/04/2021
									A070909	Oversensing	16/04/2021	16/04/2021
									A071302	Inappropriate/Inadequate Shock/Stimulation	16/04/2021	16/04/2021
10367	2020	All other reportable incidents	47267	Single-chamber implantable pacemaker, rate-responsive	CT141	Electrodes/Leads and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	08/01/2021	08/01/2021
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	08/01/2021	08/01/2021
							F12	Serious Injury/ Illness/ Impairment	A070504	Premature Discharge of Battery	08/01/2021	08/01/2021
							F1903	Device Explantation	A070504	Premature Discharge of Battery	08/01/2021	08/01/2021
					CT145	Cardiovascular devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	08/01/2021	08/01/2021
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	08/01/2021	08/01/2021
							F12	Serious Injury/ Illness/ Impairment	A070504	Premature Discharge of Battery	08/01/2021	08/01/2021
							F1903	Device Explantation	A070504	Premature Discharge of Battery	08/01/2021	08/01/2021
					CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	08/01/2021	08/01/2021
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	08/01/2021	08/01/2021
							F12	Serious Injury/ Illness/ Impairment	A070504	Premature Discharge of Battery	08/01/2021	08/01/2021
							F1903	Device Explantation	A070504	Premature Discharge of Battery	08/01/2021	08/01/2021
					CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	08/01/2021	08/01/2021
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	08/01/2021	08/01/2021
							F12	Serious Injury/ Illness/ Impairment	A070504	Premature Discharge of Battery	08/01/2021	08/01/2021
							F1903	Device Explantation	A070504	Premature Discharge of Battery	08/01/2021	08/01/2021
					CT233	Surgical	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	08/01/2021	08/01/2021
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	08/01/2021	08/01/2021
							F12	Serious Injury/ Illness/ Impairment	A070504	Premature Discharge of Battery	08/01/2021	08/01/2021
							F1903	Device Explantation	A070504	Premature Discharge of Battery	08/01/2021	08/01/2021
					CT2406	Implantable	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	08/01/2021	08/01/2021
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	08/01/2021	08/01/2021
							F12	Serious Injury/ Illness/ Impairment	A070504	Premature Discharge of Battery	08/01/2021	08/01/2021
							F1903	Device Explantation	A070504	Premature Discharge of Battery	08/01/2021	08/01/2021
					CT2407	Active implantable	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	08/01/2021	08/01/2021
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	08/01/2021	08/01/2021
							F12	Serious Injury/ Illness/ Impairment	A070504	Premature Discharge of Battery	08/01/2021	08/01/2021
							F1903	Device Explantation	A070504	Premature Discharge of Battery	08/01/2021	08/01/2021
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	08/01/2021	08/01/2021
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	08/01/2021	08/01/2021
							F12	Serious Injury/ Illness/ Impairment	A070504	Premature Discharge of Battery	08/01/2021	08/01/2021
							F1903	Device Explantation	A070504	Premature Discharge of Battery	08/01/2021	08/01/2021
					CT2727	Active implant devices/ancillaries	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	08/01/2021	08/01/2021
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	08/01/2021	08/01/2021
							F12	Serious Injury/ Illness/ Impairment	A070504	Premature Discharge of Battery	08/01/2021	08/01/2021
							F1903	Device Explantation	A070504	Premature Discharge of Battery	08/01/2021	08/01/2021
					CT2750	Implantable cardiac electrical stimulators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	08/01/2021	08/01/2021
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	08/01/2021	08/01/2021
							F12	Serious Injury/ Illness/ Impairment	A070504	Premature Discharge of Battery	08/01/2021	08/01/2021
							F1903	Device Explantation	A070504	Premature Discharge of Battery	08/01/2021	08/01/2021
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	08/01/2021	08/01/2021
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	08/01/2021	08/01/2021
							F12	Serious Injury/ Illness/ Impairment	A070504	Premature Discharge of Battery	08/01/2021	08/01/2021
							F1903	Device Explantation	A070504	Premature Discharge of Battery	08/01/2021	08/01/2021
					CT2985	Active	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	08/01/2021	08/01/2021
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	08/01/2021	08/01/2021
							F12	Serious Injury/ Illness/ Impairment	A070504	Premature Discharge of Battery	08/01/2021	08/01/2021
							F1903	Device Explantation	A070504	Premature Discharge of Battery	08/01/2021	08/01/2021
					CT3001	Implantable metal	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	08/01/2021	08/01/2021
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	08/01/2021	08/01/2021
							F12	Serious Injury/ Illness/ Impairment	A070504	Premature Discharge of Battery	08/01/2021	08/01/2021
							F1903	Device Explantation	A070504	Premature Discharge of Battery	08/01/2021	08/01/2021
					CT3002	Implantable synthetic polymer	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	08/01/2021	08/01/2021
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	08/01/2021	08/01/2021
							F12	Serious Injury/ Illness/ Impairment	A070504	Premature Discharge of Battery	08/01/2021	08/01/2021
							F1903	Device Explantation	A070504	Premature Discharge of Battery	08/01/2021	08/01/2021
					CT3006	Implantable inorganic material	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	08/01/2021	08/01/2021
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	08/01/2021	08/01/2021
							F12	Serious Injury/ Illness/ Impairment	A070504	Premature Discharge of Battery	08/01/2021	08/01/2021
							F1903	Device Explantation	A070504	Premature Discharge of Battery	08/01/2021	08/01/2021
					CT314	Electrically powered	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	08/01/2021	08/01/2021
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	08/01/2021	08/01/2021
							F12	Serious Injury/ Illness/ Impairment	A070504	Premature Discharge of Battery	08/01/2021	08/01/2021
							F1903	Device Explantation	A070504	Premature Discharge of Battery	08/01/2021	08/01/2021
					CT321	Long-term surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	08/01/2021	08/01/2021

							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	08/01/2021	08/01/2021
							F12	Serious Injury/ Illness/ Impairment	A070504	Premature Discharge of Battery	08/01/2021	08/01/2021
							F1903	Device Explantation	A070504	Premature Discharge of Battery	08/01/2021	08/01/2021
					CT449	Generators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	08/01/2021	08/01/2021
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	08/01/2021	08/01/2021
							F12	Serious Injury/ Illness/ Impairment	A070504	Premature Discharge of Battery	08/01/2021	08/01/2021
							F1903	Device Explantation	A070504	Premature Discharge of Battery	08/01/2021	08/01/2021
					CT981	Single use	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	08/01/2021	08/01/2021
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	08/01/2021	08/01/2021
							F12	Serious Injury/ Illness/ Impairment	A070504	Premature Discharge of Battery	08/01/2021	08/01/2021
							F1903	Device Explantation	A070504	Premature Discharge of Battery	08/01/2021	08/01/2021
					CT983	Surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	08/01/2021	08/01/2021
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	08/01/2021	08/01/2021
							F12	Serious Injury/ Illness/ Impairment	A070504	Premature Discharge of Battery	08/01/2021	08/01/2021
							F1903	Device Explantation	A070504	Premature Discharge of Battery	08/01/2021	08/01/2021
10368	2020	All other reportable incidents	35852	Single-chamber implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E2104	Electric Shock	A070504	Premature Discharge of Battery	16/04/2021	16/04/2021
									A070909	Oversensing	16/04/2021	16/04/2021
									A071302	Inappropriate/Inadequate Shock/Stimulation	16/04/2021	16/04/2021
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	16/04/2021	16/04/2021
									A070909	Oversensing	16/04/2021	16/04/2021
									A071302	Inappropriate/Inadequate Shock/Stimulation	16/04/2021	16/04/2021
							F14	Prolonged Episode of Care	A070504	Premature Discharge of Battery	16/04/2021	16/04/2021
									A070909	Oversensing	16/04/2021	16/04/2021
									A071302	Inappropriate/Inadequate Shock/Stimulation	16/04/2021	16/04/2021
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	16/04/2021	16/04/2021
									A070909	Oversensing	16/04/2021	16/04/2021
									A071302	Inappropriate/Inadequate Shock/Stimulation	16/04/2021	16/04/2021
10369	2020	All other reportable incidents	35852	Single-chamber implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem		
10370	2020	All other reportable incidents	35852	Single-chamber implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	03/12/2020	03/12/2020
							F1901	Additional Surgery	A070504	Premature Discharge of Battery	03/12/2020	03/12/2020
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	03/12/2020	03/12/2020
10371	2020	All other reportable incidents	47265	Dual-chamber implantable pacemaker, rate-responsive	CT141	Electrodes/Leads and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem	30/11/2020	30/11/2020
									A070504	Premature Discharge of Battery	30/11/2020	30/11/2020
							F19	Surgical Intervention	A0705	Battery Problem	30/11/2020	30/11/2020
									A070504	Premature Discharge of Battery	30/11/2020	30/11/2020
					CT145	Cardiovascular devices	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem	30/11/2020	30/11/2020
									A070504	Premature Discharge of Battery	30/11/2020	30/11/2020
							F19	Surgical Intervention	A0705	Battery Problem	30/11/2020	30/11/2020
									A070504	Premature Discharge of Battery	30/11/2020	30/11/2020
					CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem	30/11/2020	30/11/2020
									A070504	Premature Discharge of Battery	30/11/2020	30/11/2020
							F19	Surgical Intervention	A0705	Battery Problem	30/11/2020	30/11/2020
									A070504	Premature Discharge of Battery	30/11/2020	30/11/2020
					CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem	30/11/2020	30/11/2020
									A070504	Premature Discharge of Battery	30/11/2020	30/11/2020
							F19	Surgical Intervention	A0705	Battery Problem	30/11/2020	30/11/2020
									A070504	Premature Discharge of Battery	30/11/2020	30/11/2020
					CT233	Surgical	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem	30/11/2020	30/11/2020
									A070504	Premature Discharge of Battery	30/11/2020	30/11/2020
							F19	Surgical Intervention	A0705	Battery Problem	30/11/2020	30/11/2020
									A070504	Premature Discharge of Battery	30/11/2020	30/11/2020
					CT2406	Implantable	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem	30/11/2020	30/11/2020
									A070504	Premature Discharge of Battery	30/11/2020	30/11/2020
							F19	Surgical Intervention	A0705	Battery Problem	30/11/2020	30/11/2020
									A070504	Premature Discharge of Battery	30/11/2020	30/11/2020
					CT2407	Active implantable	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem	30/11/2020	30/11/2020
									A070504	Premature Discharge of Battery	30/11/2020	30/11/2020
							F19	Surgical Intervention	A0705	Battery Problem	30/11/2020	30/11/2020
									A070504	Premature Discharge of Battery	30/11/2020	30/11/2020
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem	30/11/2020	30/11/2020
									A070504	Premature Discharge of Battery	30/11/2020	30/11/2020
							F19	Surgical Intervention	A0705	Battery Problem	30/11/2020	30/11/2020
									A070504	Premature Discharge of Battery	30/11/2020	30/11/2020
					CT2727	Active implant devices/ancillaries	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem	30/11/2020	30/11/2020
									A070504	Premature Discharge of Battery	30/11/2020	30/11/2020
							F19	Surgical Intervention	A0705	Battery Problem	30/11/2020	30/11/2020
									A070504	Premature Discharge of Battery	30/11/2020	30/11/2020
					CT2750	Implantable cardiac electrical stimulators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem	30/11/2020	30/11/2020
									A070504	Premature Discharge of Battery	30/11/2020	30/11/2020
							F19	Surgical Intervention	A0705	Battery Problem	30/11/2020	30/11/2020
									A070504	Premature Discharge of Battery	30/11/2020	30/11/2020
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem	30/11/2020	30/11/2020
									A070504	Premature Discharge of Battery	30/11/2020	30/11/2020
							F19	Surgical Intervention	A0705	Battery Problem	30/11/2020	30/11/2020
									A070504	Premature Discharge of Battery	30/11/2020	30/11/2020
					CT2985	Active	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem	30/11/2020	30/11/2020
									A070504	Premature Discharge of Battery	30/11/2020	30/11/2020
							F19	Surgical Intervention	A0705	Battery Problem	30/11/2020	30/11/2020
									A070504	Premature Discharge of Battery	30/11/2020	30/11/2020
					CT3001	Implantable metal	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem	30/11/2020	30/11/2020
									A070504	Premature Discharge of Battery	30/11/2020	30/11/2020
							F19	Surgical Intervention	A0705	Battery Problem	30/11/2020	30/11/2020
									A070504	Premature Discharge of Battery	30/11/2020	30/11/2020
					CT3002	Implantable synthetic polymer	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem	30/11/2020	30/11/2020
									A070504	Premature Discharge of Battery	30/11/2020	30/11/2020
							F19	Surgical Intervention	A0705	Battery Problem	30/11/2020	30/11/2020
									A070504	Premature Discharge of Battery	30/11/2020	30/11/2020
					CT3006	Implantable inorganic material	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem	30/11/2020	30/11/2020

									A070504	Premature Discharge of Battery	30/11/2020	30/11/2020
						F19	Surgical Intervention	A0705	Battery Problem	30/11/2020	30/11/2020
								A070504	Premature Discharge of Battery	30/11/2020	30/11/2020
				CT314	Electrically powered	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem	30/11/2020	30/11/2020
								A070504	Premature Discharge of Battery	30/11/2020	30/11/2020
						F19	Surgical Intervention	A0705	Battery Problem	30/11/2020	30/11/2020
								A070504	Premature Discharge of Battery	30/11/2020	30/11/2020
				CT321	Long-term surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem	30/11/2020	30/11/2020
								A070504	Premature Discharge of Battery	30/11/2020	30/11/2020
						F19	Surgical Intervention	A0705	Battery Problem	30/11/2020	30/11/2020
								A070504	Premature Discharge of Battery	30/11/2020	30/11/2020
				CT449	Generators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem	30/11/2020	30/11/2020
								A070504	Premature Discharge of Battery	30/11/2020	30/11/2020
						F19	Surgical Intervention	A0705	Battery Problem	30/11/2020	30/11/2020
								A070504	Premature Discharge of Battery	30/11/2020	30/11/2020
				CT981	Single use	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem	30/11/2020	30/11/2020
								A070504	Premature Discharge of Battery	30/11/2020	30/11/2020
						F19	Surgical Intervention	A0705	Battery Problem	30/11/2020	30/11/2020
								A070504	Premature Discharge of Battery	30/11/2020	30/11/2020
				CT983	Surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem	30/11/2020	30/11/2020
								A070504	Premature Discharge of Battery	30/11/2020	30/11/2020
						F19	Surgical Intervention	A0705	Battery Problem	30/11/2020	30/11/2020
								A070504	Premature Discharge of Battery	30/11/2020	30/11/2020
10372	2020	All other reportable incidents	35852	Single-chamber implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E2104	Electric Shock	A0712	Pacing Problem		
									A071302	Inappropriate/Inadequate Shock/Stimulation		
							F15	Recognized Device or Procedural Complication	A0712	Pacing Problem		
									A071302	Inappropriate/Inadequate Shock/Stimulation		
							F24	Insufficient Health Impact Information	A0712	Pacing Problem		
									A071302	Inappropriate/Inadequate Shock/Stimulation		
10373	2020	All other reportable incidents	37265	Dual-chamber implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	04/12/2020	04/12/2020
							F1901	Additional Surgery	A070504	Premature Discharge of Battery	04/12/2020	04/12/2020
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	04/12/2020	04/12/2020
10374	2020	All other reportable incidents	47265	Dual-chamber implantable pacemaker, rate-responsive	CT141	Electrodes/Leads and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	30/03/2019	30/03/2019
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	30/03/2019	30/03/2019
							F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A070504	Premature Discharge of Battery	30/03/2019	30/03/2019
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	30/03/2019	30/03/2019
				CT145	Cardiovascular devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	30/03/2019	30/03/2019
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	30/03/2019	30/03/2019
							F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A070504	Premature Discharge of Battery	30/03/2019	30/03/2019
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	30/03/2019	30/03/2019
				CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	30/03/2019	30/03/2019
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	30/03/2019	30/03/2019
							F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A070504	Premature Discharge of Battery	30/03/2019	30/03/2019
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	30/03/2019	30/03/2019
				CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	30/03/2019	30/03/2019
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	30/03/2019	30/03/2019
							F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A070504	Premature Discharge of Battery	30/03/2019	30/03/2019
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	30/03/2019	30/03/2019
				CT233	Surgical	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	30/03/2019	30/03/2019
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	30/03/2019	30/03/2019
							F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A070504	Premature Discharge of Battery	30/03/2019	30/03/2019
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	30/03/2019	30/03/2019
				CT2406	Implantable	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	30/03/2019	30/03/2019
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	30/03/2019	30/03/2019
							F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A070504	Premature Discharge of Battery	30/03/2019	30/03/2019
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	30/03/2019	30/03/2019
				CT2407	Active implantable	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	30/03/2019	30/03/2019
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	30/03/2019	30/03/2019
							F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A070504	Premature Discharge of Battery	30/03/2019	30/03/2019
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	30/03/2019	30/03/2019
				CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	30/03/2019	30/03/2019
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	30/03/2019	30/03/2019
							F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A070504	Premature Discharge of Battery	30/03/2019	30/03/2019
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	30/03/2019	30/03/2019
				CT2727	Active implant devices/ancillaries	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	30/03/2019	30/03/2019
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	30/03/2019	30/03/2019
							F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A070504	Premature Discharge of Battery	30/03/2019	30/03/2019
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	30/03/2019	30/03/2019
				CT2750	Implantable cardiac electrical stimulators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	30/03/2019	30/03/2019
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	30/03/2019	30/03/2019
							F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A070504	Premature Discharge of Battery	30/03/2019	30/03/2019
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	30/03/2019	30/03/2019
				CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	30/03/2019	30/03/2019
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	30/03/2019	30/03/2019
							F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A070504	Premature Discharge of Battery	30/03/2019	30/03/2019
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	30/03/2019	30/03/2019
				CT2985	Active	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	30/03/2019	30/03/2019
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	30/03/2019	30/03/2019
							F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A070504	Premature Discharge of Battery	30/03/2019	30/03/2019
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	30/03/2019	30/03/2019
				CT3001	Implantable metal	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	30/03/2019	30/03/2019
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	30/03/2019	30/03/2019
							F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A070504	Premature Discharge of Battery	30/03/2019	30/03/2019
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	30/03/2019	30/03/2019
				CT3002	Implantable synthetic polymer	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	30/03/2019	30/03/2019
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	30/03/2019	30/03/2019
							F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A070504	Premature Discharge of Battery	30/03/2019	30/03/2019
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	30/03/2019	30/03/2019

					CT3006	Implantable inorganic material	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	30/03/2019	30/03/2019
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	30/03/2019	30/03/2019
							F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A070504	Premature Discharge of Battery	30/03/2019	30/03/2019
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	30/03/2019	30/03/2019
					CT314	Electrically powered	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	30/03/2019	30/03/2019
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	30/03/2019	30/03/2019
							F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A070504	Premature Discharge of Battery	30/03/2019	30/03/2019
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	30/03/2019	30/03/2019
					CT321	Long-term surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	30/03/2019	30/03/2019
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	30/03/2019	30/03/2019
							F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A070504	Premature Discharge of Battery	30/03/2019	30/03/2019
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	30/03/2019	30/03/2019
					CT449	Generators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	30/03/2019	30/03/2019
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	30/03/2019	30/03/2019
							F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A070504	Premature Discharge of Battery	30/03/2019	30/03/2019
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	30/03/2019	30/03/2019
					CT981	Single use	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	30/03/2019	30/03/2019
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	30/03/2019	30/03/2019
							F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A070504	Premature Discharge of Battery	30/03/2019	30/03/2019
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	30/03/2019	30/03/2019
					CT983	Surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	30/03/2019	30/03/2019
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	30/03/2019	30/03/2019
							F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A070504	Premature Discharge of Battery	30/03/2019	30/03/2019
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	30/03/2019	30/03/2019
10375	2020	All other reportable incidents	47270	Cardiac resynchronization therapy implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E2104	Electric Shock	A071302	Inappropriate/Inadequate Shock/Stimulation		
							F26	No Health Consequences or Impact	A071302	Inappropriate/Inadequate Shock/Stimulation		
10376	2020	All other reportable incidents	35852	Single-chamber implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	28/01/2021	28/01/2021
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	28/01/2021	28/01/2021
							F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A070504	Premature Discharge of Battery	28/01/2021	28/01/2021
							F19	Surgical Intervention	A070504	Premature Discharge of Battery	28/01/2021	28/01/2021
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	28/01/2021	28/01/2021
10377	2020	All other reportable incidents	37265	Dual-chamber implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A070909	Oversensing		
							F26	No Health Consequences or Impact	A070909	Oversensing		
10378	2020	All other reportable incidents	17882	Physiologic-monitoring defibrillation system	CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A07	Electrical /Electronic Property Problem		
							F27	Problem Identified During Non-Clinical Procedure	A07	Electrical /Electronic Property Problem		
10379	2020	All other reportable incidents	35223	Endocardial/interventricular septal pacing lead	CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A0401	Break		
									A072202	Low Impedance		
							F1901	Additional Surgery	A0401	Break		
									A072202	Low Impedance		
							F1905	Device Revision or Replacement	A0401	Break		
									A072202	Low Impedance		
10380	2020	Death	37265	Dual-chamber implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E0601	Arrhythmia	A070910	Undersensing		
									A071301	Failure to Deliver Shock/Stimulation		
									A071302	Inappropriate/Inadequate Shock/Stimulation		
							F02	Death	A070910	Undersensing		
									A071301	Failure to Deliver Shock/Stimulation		
									A071302	Inappropriate/Inadequate Shock/Stimulation		
							F21	Unexpected Deterioration	A070910	Undersensing		
									A071301	Failure to Deliver Shock/Stimulation		
									A071302	Inappropriate/Inadequate Shock/Stimulation		
10381	2020	All other reportable incidents	37265	Dual-chamber implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A071102	Failure to Interrogate	04/12/2020	04/12/2020
							F1901	Additional Surgery	A071102	Failure to Interrogate	04/12/2020	04/12/2020
							F1905	Device Revision or Replacement	A071102	Failure to Interrogate	04/12/2020	04/12/2020
10382	2020	All other reportable incidents	36241	Pacemaker/defibrillator lead	CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070909	Oversensing		
									A072201	High Impedance		
									A12	Connection Problem		
							F26	No Health Consequences or Impact	A070909	Oversensing		
									A072201	High Impedance		
									A12	Connection Problem		
10383	2020	All other reportable incidents	36241	Pacemaker/defibrillator lead	CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A072201	High Impedance		
							F26	No Health Consequences or Impact	A072201	High Impedance		
10384	2020	All other reportable incidents	47270	Cardiac resynchronization therapy implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E2104	Electric Shock	A071302	Inappropriate/Inadequate Shock/Stimulation		
							E2403	No Clinical Signs, Symptoms or Conditions	A071302	Inappropriate/Inadequate Shock/Stimulation		
							F15	Recognized Device or Procedural Complication	A071302	Inappropriate/Inadequate Shock/Stimulation		
10385	2020	All other reportable incidents	47267	Single-chamber implantable pacemaker, rate-responsive	CT141	Electrodes/Leads and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	28/03/2021	28/03/2021
									A070909	Oversensing	28/03/2021	28/03/2021
									A072201	High Impedance	28/03/2021	28/03/2021
									A12	Connection Problem	28/03/2021	28/03/2021
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	28/03/2021	28/03/2021
									A070909	Oversensing	28/03/2021	28/03/2021
									A072201	High Impedance	28/03/2021	28/03/2021
									A12	Connection Problem	28/03/2021	28/03/2021
							F12	Serious Injury/ Illness/ Impairment	A070504	Premature Discharge of Battery	28/03/2021	28/03/2021
									A070909	Oversensing	28/03/2021	28/03/2021
									A072201	High Impedance	28/03/2021	28/03/2021
									A12	Connection Problem	28/03/2021	28/03/2021
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	28/03/2021	28/03/2021
									A070909	Oversensing	28/03/2021	28/03/2021
									A072201	High Impedance	28/03/2021	28/03/2021
									A12	Connection Problem	28/03/2021	28/03/2021
					CT145	Cardiovascular devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	28/03/2021	28/03/2021
									A070909	Oversensing	28/03/2021	28/03/2021
									A072201	High Impedance	28/03/2021	28/03/2021
									A12	Connection Problem	28/03/2021	28/03/2021
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	28/03/2021	28/03/2021
									A070909	Oversensing	28/03/2021	28/03/2021
									A072201	High Impedance	28/03/2021	28/03/2021
									A12	Connection Problem	28/03/2021	28/03/2021

						F12	Serious Injury/ Illness/ Impairment	A070504	Premature Discharge of Battery	28/03/2021	28/03/2021
								A070909	Oversensing	28/03/2021	28/03/2021
								A072201	High Impedance	28/03/2021	28/03/2021
								A12	Connection Problem	28/03/2021	28/03/2021
						F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	28/03/2021	28/03/2021
								A070909	Oversensing	28/03/2021	28/03/2021
								A072201	High Impedance	28/03/2021	28/03/2021
								A12	Connection Problem	28/03/2021	28/03/2021
					CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	28/03/2021	28/03/2021
									A070909	Oversensing	28/03/2021	28/03/2021
									A072201	High Impedance	28/03/2021	28/03/2021
									A12	Connection Problem	28/03/2021	28/03/2021
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	28/03/2021	28/03/2021
									A070909	Oversensing	28/03/2021	28/03/2021
									A072201	High Impedance	28/03/2021	28/03/2021
									A12	Connection Problem	28/03/2021	28/03/2021
							F12	Serious Injury/ Illness/ Impairment	A070504	Premature Discharge of Battery	28/03/2021	28/03/2021
									A070909	Oversensing	28/03/2021	28/03/2021
									A072201	High Impedance	28/03/2021	28/03/2021
									A12	Connection Problem	28/03/2021	28/03/2021
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	28/03/2021	28/03/2021
									A070909	Oversensing	28/03/2021	28/03/2021
									A072201	High Impedance	28/03/2021	28/03/2021
									A12	Connection Problem	28/03/2021	28/03/2021
					CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	28/03/2021	28/03/2021
									A070909	Oversensing	28/03/2021	28/03/2021
									A072201	High Impedance	28/03/2021	28/03/2021
									A12	Connection Problem	28/03/2021	28/03/2021
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	28/03/2021	28/03/2021
									A070909	Oversensing	28/03/2021	28/03/2021
									A072201	High Impedance	28/03/2021	28/03/2021
									A12	Connection Problem	28/03/2021	28/03/2021
							F12	Serious Injury/ Illness/ Impairment	A070504	Premature Discharge of Battery	28/03/2021	28/03/2021
									A070909	Oversensing	28/03/2021	28/03/2021
									A072201	High Impedance	28/03/2021	28/03/2021
									A12	Connection Problem	28/03/2021	28/03/2021
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	28/03/2021	28/03/2021
									A070909	Oversensing	28/03/2021	28/03/2021
									A072201	High Impedance	28/03/2021	28/03/2021
									A12	Connection Problem	28/03/2021	28/03/2021
					CT233	Surgical	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	28/03/2021	28/03/2021
									A070909	Oversensing	28/03/2021	28/03/2021
									A072201	High Impedance	28/03/2021	28/03/2021
									A12	Connection Problem	28/03/2021	28/03/2021
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	28/03/2021	28/03/2021
									A070909	Oversensing	28/03/2021	28/03/2021
									A072201	High Impedance	28/03/2021	28/03/2021
									A12	Connection Problem	28/03/2021	28/03/2021
							F12	Serious Injury/ Illness/ Impairment	A070504	Premature Discharge of Battery	28/03/2021	28/03/2021
									A070909	Oversensing	28/03/2021	28/03/2021
									A072201	High Impedance	28/03/2021	28/03/2021
									A12	Connection Problem	28/03/2021	28/03/2021
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	28/03/2021	28/03/2021
									A070909	Oversensing	28/03/2021	28/03/2021
									A072201	High Impedance	28/03/2021	28/03/2021
									A12	Connection Problem	28/03/2021	28/03/2021
					CT2406	Implantable	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	28/03/2021	28/03/2021
									A070909	Oversensing	28/03/2021	28/03/2021
									A072201	High Impedance	28/03/2021	28/03/2021
									A12	Connection Problem	28/03/2021	28/03/2021
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	28/03/2021	28/03/2021
									A070909	Oversensing	28/03/2021	28/03/2021
									A072201	High Impedance	28/03/2021	28/03/2021
									A12	Connection Problem	28/03/2021	28/03/2021
							F12	Serious Injury/ Illness/ Impairment	A070504	Premature Discharge of Battery	28/03/2021	28/03/2021
									A070909	Oversensing	28/03/2021	28/03/2021
									A072201	High Impedance	28/03/2021	28/03/2021
									A12	Connection Problem	28/03/2021	28/03/2021
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	28/03/2021	28/03/2021
									A070909	Oversensing	28/03/2021	28/03/2021
									A072201	High Impedance	28/03/2021	28/03/2021
									A12	Connection Problem	28/03/2021	28/03/2021
					CT2407	Active implantable	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	28/03/2021	28/03/2021
									A070909	Oversensing	28/03/2021	28/03/2021
									A072201	High Impedance	28/03/2021	28/03/2021
									A12	Connection Problem	28/03/2021	28/03/2021
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	28/03/2021	28/03/2021
									A070909	Oversensing	28/03/2021	28/03/2021
									A072201	High Impedance	28/03/2021	28/03/2021
									A12	Connection Problem	28/03/2021	28/03/2021
							F12	Serious Injury/ Illness/ Impairment	A070504	Premature Discharge of Battery	28/03/2021	28/03/2021
									A070909	Oversensing	28/03/2021	28/03/2021
									A072201	High Impedance	28/03/2021	28/03/2021
									A12	Connection Problem	28/03/2021	28/03/2021
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	28/03/2021	28/03/2021
									A070909	Oversensing	28/03/2021	28/03/2021
									A072201	High Impedance	28/03/2021	28/03/2021
									A12	Connection Problem	28/03/2021	28/03/2021

					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	28/03/2021	28/03/2021
									A070909	Oversensing	28/03/2021	28/03/2021
									A072201	High Impedance	28/03/2021	28/03/2021
									A12	Connection Problem	28/03/2021	28/03/2021
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	28/03/2021	28/03/2021
									A070909	Oversensing	28/03/2021	28/03/2021
									A072201	High Impedance	28/03/2021	28/03/2021
									A12	Connection Problem	28/03/2021	28/03/2021
							F12	Serious Injury/ Illness/ Impairment	A070504	Premature Discharge of Battery	28/03/2021	28/03/2021
									A070909	Oversensing	28/03/2021	28/03/2021
									A072201	High Impedance	28/03/2021	28/03/2021
									A12	Connection Problem	28/03/2021	28/03/2021
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	28/03/2021	28/03/2021
									A070909	Oversensing	28/03/2021	28/03/2021
									A072201	High Impedance	28/03/2021	28/03/2021
									A12	Connection Problem	28/03/2021	28/03/2021
					CT2727	Active implant devices/ancillaries	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	28/03/2021	28/03/2021
									A070909	Oversensing	28/03/2021	28/03/2021
									A072201	High Impedance	28/03/2021	28/03/2021
									A12	Connection Problem	28/03/2021	28/03/2021
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	28/03/2021	28/03/2021
									A070909	Oversensing	28/03/2021	28/03/2021
									A072201	High Impedance	28/03/2021	28/03/2021
									A12	Connection Problem	28/03/2021	28/03/2021
							F12	Serious Injury/ Illness/ Impairment	A070504	Premature Discharge of Battery	28/03/2021	28/03/2021
									A070909	Oversensing	28/03/2021	28/03/2021
									A072201	High Impedance	28/03/2021	28/03/2021
									A12	Connection Problem	28/03/2021	28/03/2021
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	28/03/2021	28/03/2021
									A070909	Oversensing	28/03/2021	28/03/2021
									A072201	High Impedance	28/03/2021	28/03/2021
									A12	Connection Problem	28/03/2021	28/03/2021
					CT2750	Implantable cardiac electrical stimulators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	28/03/2021	28/03/2021
									A070909	Oversensing	28/03/2021	28/03/2021
									A072201	High Impedance	28/03/2021	28/03/2021
									A12	Connection Problem	28/03/2021	28/03/2021
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	28/03/2021	28/03/2021
									A070909	Oversensing	28/03/2021	28/03/2021
									A072201	High Impedance	28/03/2021	28/03/2021
									A12	Connection Problem	28/03/2021	28/03/2021
							F12	Serious Injury/ Illness/ Impairment	A070504	Premature Discharge of Battery	28/03/2021	28/03/2021
									A070909	Oversensing	28/03/2021	28/03/2021
									A072201	High Impedance	28/03/2021	28/03/2021
									A12	Connection Problem	28/03/2021	28/03/2021
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	28/03/2021	28/03/2021
									A070909	Oversensing	28/03/2021	28/03/2021
									A072201	High Impedance	28/03/2021	28/03/2021
									A12	Connection Problem	28/03/2021	28/03/2021
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	28/03/2021	28/03/2021
									A070909	Oversensing	28/03/2021	28/03/2021
									A072201	High Impedance	28/03/2021	28/03/2021
									A12	Connection Problem	28/03/2021	28/03/2021
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	28/03/2021	28/03/2021
									A070909	Oversensing	28/03/2021	28/03/2021
									A072201	High Impedance	28/03/2021	28/03/2021
									A12	Connection Problem	28/03/2021	28/03/2021
							F12	Serious Injury/ Illness/ Impairment	A070504	Premature Discharge of Battery	28/03/2021	28/03/2021
									A070909	Oversensing	28/03/2021	28/03/2021
									A072201	High Impedance	28/03/2021	28/03/2021
									A12	Connection Problem	28/03/2021	28/03/2021
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	28/03/2021	28/03/2021
									A070909	Oversensing	28/03/2021	28/03/2021
									A072201	High Impedance	28/03/2021	28/03/2021
									A12	Connection Problem	28/03/2021	28/03/2021
					CT2985	Active	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	28/03/2021	28/03/2021
									A070909	Oversensing	28/03/2021	28/03/2021
									A072201	High Impedance	28/03/2021	28/03/2021
									A12	Connection Problem	28/03/2021	28/03/2021
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	28/03/2021	28/03/2021
									A070909	Oversensing	28/03/2021	28/03/2021
									A072201	High Impedance	28/03/2021	28/03/2021
									A12	Connection Problem	28/03/2021	28/03/2021
							F12	Serious Injury/ Illness/ Impairment	A070504	Premature Discharge of Battery	28/03/2021	28/03/2021
									A070909	Oversensing	28/03/2021	28/03/2021
									A072201	High Impedance	28/03/2021	28/03/2021
									A12	Connection Problem	28/03/2021	28/03/2021
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	28/03/2021	28/03/2021
									A070909	Oversensing	28/03/2021	28/03/2021
									A072201	High Impedance	28/03/2021	28/03/2021
									A12	Connection Problem	28/03/2021	28/03/2021
					CT3001	Implantable metal	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	28/03/2021	28/03/2021
									A070909	Oversensing	28/03/2021	28/03/2021
									A072201	High Impedance	28/03/2021	28/03/2021
									A12	Connection Problem	28/03/2021	28/03/2021
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	28/03/2021	28/03/2021
									A070909	Oversensing	28/03/2021	28/03/2021
									A072201	High Impedance	28/03/2021	28/03/2021
									A12	Connection Problem	28/03/2021	28/03/2021

						F12	Serious Injury/ Illness/ Impairment	A070504	Premature Discharge of Battery	28/03/2021	28/03/2021
								A070909	Oversensing	28/03/2021	28/03/2021
								A072201	High Impedance	28/03/2021	28/03/2021
								A12	Connection Problem	28/03/2021	28/03/2021
						F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	28/03/2021	28/03/2021
								A070909	Oversensing	28/03/2021	28/03/2021
								A072201	High Impedance	28/03/2021	28/03/2021
								A12	Connection Problem	28/03/2021	28/03/2021
					CT3002	Implantable synthetic polymer	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	28/03/2021	28/03/2021
									A070909	Oversensing	28/03/2021	28/03/2021
									A072201	High Impedance	28/03/2021	28/03/2021
									A12	Connection Problem	28/03/2021	28/03/2021
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	28/03/2021	28/03/2021
									A070909	Oversensing	28/03/2021	28/03/2021
									A072201	High Impedance	28/03/2021	28/03/2021
									A12	Connection Problem	28/03/2021	28/03/2021
							F12	Serious Injury/ Illness/ Impairment	A070504	Premature Discharge of Battery	28/03/2021	28/03/2021
									A070909	Oversensing	28/03/2021	28/03/2021
									A072201	High Impedance	28/03/2021	28/03/2021
									A12	Connection Problem	28/03/2021	28/03/2021
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	28/03/2021	28/03/2021
									A070909	Oversensing	28/03/2021	28/03/2021
									A072201	High Impedance	28/03/2021	28/03/2021
									A12	Connection Problem	28/03/2021	28/03/2021
					CT3006	Implantable inorganic material	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	28/03/2021	28/03/2021
									A070909	Oversensing	28/03/2021	28/03/2021
									A072201	High Impedance	28/03/2021	28/03/2021
									A12	Connection Problem	28/03/2021	28/03/2021
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	28/03/2021	28/03/2021
									A070909	Oversensing	28/03/2021	28/03/2021
									A072201	High Impedance	28/03/2021	28/03/2021
									A12	Connection Problem	28/03/2021	28/03/2021
							F12	Serious Injury/ Illness/ Impairment	A070504	Premature Discharge of Battery	28/03/2021	28/03/2021
									A070909	Oversensing	28/03/2021	28/03/2021
									A072201	High Impedance	28/03/2021	28/03/2021
									A12	Connection Problem	28/03/2021	28/03/2021
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	28/03/2021	28/03/2021
									A070909	Oversensing	28/03/2021	28/03/2021
									A072201	High Impedance	28/03/2021	28/03/2021
									A12	Connection Problem	28/03/2021	28/03/2021
					CT314	Electrically powered	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	28/03/2021	28/03/2021
									A070909	Oversensing	28/03/2021	28/03/2021
									A072201	High Impedance	28/03/2021	28/03/2021
									A12	Connection Problem	28/03/2021	28/03/2021
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	28/03/2021	28/03/2021
									A070909	Oversensing	28/03/2021	28/03/2021
									A072201	High Impedance	28/03/2021	28/03/2021
									A12	Connection Problem	28/03/2021	28/03/2021
							F12	Serious Injury/ Illness/ Impairment	A070504	Premature Discharge of Battery	28/03/2021	28/03/2021
									A070909	Oversensing	28/03/2021	28/03/2021
									A072201	High Impedance	28/03/2021	28/03/2021
									A12	Connection Problem	28/03/2021	28/03/2021
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	28/03/2021	28/03/2021
									A070909	Oversensing	28/03/2021	28/03/2021
									A072201	High Impedance	28/03/2021	28/03/2021
									A12	Connection Problem	28/03/2021	28/03/2021
					CT321	Long-term surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	28/03/2021	28/03/2021
									A070909	Oversensing	28/03/2021	28/03/2021
									A072201	High Impedance	28/03/2021	28/03/2021
									A12	Connection Problem	28/03/2021	28/03/2021
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	28/03/2021	28/03/2021
									A070909	Oversensing	28/03/2021	28/03/2021
									A072201	High Impedance	28/03/2021	28/03/2021
									A12	Connection Problem	28/03/2021	28/03/2021
							F12	Serious Injury/ Illness/ Impairment	A070504	Premature Discharge of Battery	28/03/2021	28/03/2021
									A070909	Oversensing	28/03/2021	28/03/2021
									A072201	High Impedance	28/03/2021	28/03/2021
									A12	Connection Problem	28/03/2021	28/03/2021
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	28/03/2021	28/03/2021
									A070909	Oversensing	28/03/2021	28/03/2021
									A072201	High Impedance	28/03/2021	28/03/2021
									A12	Connection Problem	28/03/2021	28/03/2021
					CT449	Generators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	28/03/2021	28/03/2021
									A070909	Oversensing	28/03/2021	28/03/2021
									A072201	High Impedance	28/03/2021	28/03/2021
									A12	Connection Problem	28/03/2021	28/03/2021
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	28/03/2021	28/03/2021
									A070909	Oversensing	28/03/2021	28/03/2021
									A072201	High Impedance	28/03/2021	28/03/2021
									A12	Connection Problem	28/03/2021	28/03/2021
							F12	Serious Injury/ Illness/ Impairment	A070504	Premature Discharge of Battery	28/03/2021	28/03/2021
									A070909	Oversensing	28/03/2021	28/03/2021
									A072201	High Impedance	28/03/2021	28/03/2021
									A12	Connection Problem	28/03/2021	28/03/2021
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	28/03/2021	28/03/2021
									A070909	Oversensing	28/03/2021	28/03/2021
									A072201	High Impedance	28/03/2021	28/03/2021
									A12	Connection Problem	28/03/2021	28/03/2021

					CT981	Single use	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	28/03/2021	28/03/2021
									A070909	Oversensing	28/03/2021	28/03/2021
									A072201	High Impedance	28/03/2021	28/03/2021
									A12	Connection Problem	28/03/2021	28/03/2021
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	28/03/2021	28/03/2021
									A070909	Oversensing	28/03/2021	28/03/2021
									A072201	High Impedance	28/03/2021	28/03/2021
									A12	Connection Problem	28/03/2021	28/03/2021
							F12	Serious Injury/ Illness/ Impairment	A070504	Premature Discharge of Battery	28/03/2021	28/03/2021
									A070909	Oversensing	28/03/2021	28/03/2021
									A072201	High Impedance	28/03/2021	28/03/2021
									A12	Connection Problem	28/03/2021	28/03/2021
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	28/03/2021	28/03/2021
									A070909	Oversensing	28/03/2021	28/03/2021
									A072201	High Impedance	28/03/2021	28/03/2021
									A12	Connection Problem	28/03/2021	28/03/2021
					CT983	Surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	28/03/2021	28/03/2021
									A070909	Oversensing	28/03/2021	28/03/2021
									A072201	High Impedance	28/03/2021	28/03/2021
									A12	Connection Problem	28/03/2021	28/03/2021
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	28/03/2021	28/03/2021
									A070909	Oversensing	28/03/2021	28/03/2021
									A072201	High Impedance	28/03/2021	28/03/2021
									A12	Connection Problem	28/03/2021	28/03/2021
							F12	Serious Injury/ Illness/ Impairment	A070504	Premature Discharge of Battery	28/03/2021	28/03/2021
									A070909	Oversensing	28/03/2021	28/03/2021
									A072201	High Impedance	28/03/2021	28/03/2021
									A12	Connection Problem	28/03/2021	28/03/2021
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	28/03/2021	28/03/2021
									A070909	Oversensing	28/03/2021	28/03/2021
									A072201	High Impedance	28/03/2021	28/03/2021
									A12	Connection Problem	28/03/2021	28/03/2021
10386	2020	All other reportable incidents	35852	Single-chamber implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A070909	Oversensing	20/11/2020	20/11/2020
									A071302	Inappropriate/Inadequate Shock/Stimulation	20/11/2020	20/11/2020
									A090801	Signal Artifact/Noise	20/11/2020	20/11/2020
							F08	Hospitalization or Prolonged Hospitalization	A070909	Oversensing	20/11/2020	20/11/2020
									A071302	Inappropriate/Inadequate Shock/Stimulation	20/11/2020	20/11/2020
									A090801	Signal Artifact/Noise	20/11/2020	20/11/2020
							F12	Serious Injury/ Illness/ Impairment	A070909	Oversensing	20/11/2020	20/11/2020
									A071302	Inappropriate/Inadequate Shock/Stimulation	20/11/2020	20/11/2020
									A090801	Signal Artifact/Noise	20/11/2020	20/11/2020
							F19	Surgical Intervention	A070909	Oversensing	20/11/2020	20/11/2020
									A071302	Inappropriate/Inadequate Shock/Stimulation	20/11/2020	20/11/2020
									A090801	Signal Artifact/Noise	20/11/2020	20/11/2020
							F1905	Device Revision or Replacement	A070909	Oversensing	20/11/2020	20/11/2020
									A071302	Inappropriate/Inadequate Shock/Stimulation	20/11/2020	20/11/2020
									A090801	Signal Artifact/Noise	20/11/2020	20/11/2020
10387	2020	All other reportable incidents	35223	Endocardial/interventricular septal pacing lead	CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A0401	Break		
									A070909	Oversensing		
							F26	No Health Consequences or Impact	A0401	Break		
									A070909	Oversensing		
10388	2020	All other reportable incidents	35223	Endocardial/interventricular septal pacing lead	CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A0401	Break		
									A070901	Decreased Sensitivity		
									A072202	Low Impedance		
							F26	No Health Consequences or Impact	A0401	Break		
									A070901	Decreased Sensitivity		
									A072202	Low Impedance		
10389	2020	All other reportable incidents	36241	Pacemaker/defibrillator lead	CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070909	Oversensing		
									A071302	Inappropriate/Inadequate Shock/Stimulation		
							F19	Surgical Intervention	A070909	Oversensing		
									A071302	Inappropriate/Inadequate Shock/Stimulation		
10390	2020	All other reportable incidents	35852	Single-chamber implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A070909	Oversensing		
									A0712	Pacing Problem		
							F26	No Health Consequences or Impact	A070909	Oversensing		
									A0712	Pacing Problem		
10391	2020	All other reportable incidents	36241	Pacemaker/defibrillator lead	CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070101	Failure to Capture		
									A0722	Impedance Problem		
							F26	No Health Consequences or Impact	A070101	Failure to Capture		
									A0722	Impedance Problem		
10392	2020	All other reportable incidents	47265	Dual-chamber implantable pacemaker, rate-responsive	CT141	Electrodes/Leads and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	25/01/2021	25/01/2021
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	25/01/2021	25/01/2021
							F12	Serious Injury/ Illness/ Impairment	A070504	Premature Discharge of Battery	25/01/2021	25/01/2021
							F19	Surgical Intervention	A070504	Premature Discharge of Battery	25/01/2021	25/01/2021
					CT145	Cardiovascular devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	25/01/2021	25/01/2021
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	25/01/2021	25/01/2021
							F12	Serious Injury/ Illness/ Impairment	A070504	Premature Discharge of Battery	25/01/2021	25/01/2021
							F19	Surgical Intervention	A070504	Premature Discharge of Battery	25/01/2021	25/01/2021
					CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	25/01/2021	25/01/2021
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	25/01/2021	25/01/2021
							F12	Serious Injury/ Illness/ Impairment	A070504	Premature Discharge of Battery	25/01/2021	25/01/2021
							F19	Surgical Intervention	A070504	Premature Discharge of Battery	25/01/2021	25/01/2021
					CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	25/01/2021	25/01/2021
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	25/01/2021	25/01/2021
							F12	Serious Injury/ Illness/ Impairment	A070504	Premature Discharge of Battery	25/01/2021	25/01/2021
							F19	Surgical Intervention	A070504	Premature Discharge of Battery	25/01/2021	25/01/2021
					CT233	Surgical	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	25/01/2021	25/01/2021
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	25/01/2021	25/01/2021
							F12	Serious Injury/ Illness/ Impairment	A070504	Premature Discharge of Battery	25/01/2021	25/01/2021

							F19	Surgical Intervention	A070504	Premature Discharge of Battery	25/01/2021	25/01/2021
					CT2406	Implantable	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	25/01/2021	25/01/2021
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	25/01/2021	25/01/2021
							F12	Serious Injury/ Illness/ Impairment	A070504	Premature Discharge of Battery	25/01/2021	25/01/2021
							F19	Surgical Intervention	A070504	Premature Discharge of Battery	25/01/2021	25/01/2021
					CT2407	Active implantable	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	25/01/2021	25/01/2021
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	25/01/2021	25/01/2021
							F12	Serious Injury/ Illness/ Impairment	A070504	Premature Discharge of Battery	25/01/2021	25/01/2021
							F19	Surgical Intervention	A070504	Premature Discharge of Battery	25/01/2021	25/01/2021
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	25/01/2021	25/01/2021
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	25/01/2021	25/01/2021
							F12	Serious Injury/ Illness/ Impairment	A070504	Premature Discharge of Battery	25/01/2021	25/01/2021
							F19	Surgical Intervention	A070504	Premature Discharge of Battery	25/01/2021	25/01/2021
					CT2727	Active implant devices/ancillaries	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	25/01/2021	25/01/2021
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	25/01/2021	25/01/2021
							F12	Serious Injury/ Illness/ Impairment	A070504	Premature Discharge of Battery	25/01/2021	25/01/2021
							F19	Surgical Intervention	A070504	Premature Discharge of Battery	25/01/2021	25/01/2021
					CT2750	Implantable cardiac electrical stimulators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	25/01/2021	25/01/2021
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	25/01/2021	25/01/2021
							F12	Serious Injury/ Illness/ Impairment	A070504	Premature Discharge of Battery	25/01/2021	25/01/2021
							F19	Surgical Intervention	A070504	Premature Discharge of Battery	25/01/2021	25/01/2021
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	25/01/2021	25/01/2021
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	25/01/2021	25/01/2021
							F12	Serious Injury/ Illness/ Impairment	A070504	Premature Discharge of Battery	25/01/2021	25/01/2021
							F19	Surgical Intervention	A070504	Premature Discharge of Battery	25/01/2021	25/01/2021
					CT2985	Active	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	25/01/2021	25/01/2021
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	25/01/2021	25/01/2021
							F12	Serious Injury/ Illness/ Impairment	A070504	Premature Discharge of Battery	25/01/2021	25/01/2021
							F19	Surgical Intervention	A070504	Premature Discharge of Battery	25/01/2021	25/01/2021
					CT3001	Implantable metal	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	25/01/2021	25/01/2021
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	25/01/2021	25/01/2021
							F12	Serious Injury/ Illness/ Impairment	A070504	Premature Discharge of Battery	25/01/2021	25/01/2021
							F19	Surgical Intervention	A070504	Premature Discharge of Battery	25/01/2021	25/01/2021
					CT3002	Implantable synthetic polymer	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	25/01/2021	25/01/2021
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	25/01/2021	25/01/2021
							F12	Serious Injury/ Illness/ Impairment	A070504	Premature Discharge of Battery	25/01/2021	25/01/2021
							F19	Surgical Intervention	A070504	Premature Discharge of Battery	25/01/2021	25/01/2021
					CT3006	Implantable inorganic material	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	25/01/2021	25/01/2021
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	25/01/2021	25/01/2021
							F12	Serious Injury/ Illness/ Impairment	A070504	Premature Discharge of Battery	25/01/2021	25/01/2021
							F19	Surgical Intervention	A070504	Premature Discharge of Battery	25/01/2021	25/01/2021
					CT314	Electrically powered	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	25/01/2021	25/01/2021
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	25/01/2021	25/01/2021
							F12	Serious Injury/ Illness/ Impairment	A070504	Premature Discharge of Battery	25/01/2021	25/01/2021
							F19	Surgical Intervention	A070504	Premature Discharge of Battery	25/01/2021	25/01/2021
					CT321	Long-term surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	25/01/2021	25/01/2021
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	25/01/2021	25/01/2021
							F12	Serious Injury/ Illness/ Impairment	A070504	Premature Discharge of Battery	25/01/2021	25/01/2021
							F19	Surgical Intervention	A070504	Premature Discharge of Battery	25/01/2021	25/01/2021
					CT449	Generators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	25/01/2021	25/01/2021
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	25/01/2021	25/01/2021
							F12	Serious Injury/ Illness/ Impairment	A070504	Premature Discharge of Battery	25/01/2021	25/01/2021
							F19	Surgical Intervention	A070504	Premature Discharge of Battery	25/01/2021	25/01/2021
					CT981	Single use	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	25/01/2021	25/01/2021
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	25/01/2021	25/01/2021
							F12	Serious Injury/ Illness/ Impairment	A070504	Premature Discharge of Battery	25/01/2021	25/01/2021
							F19	Surgical Intervention	A070504	Premature Discharge of Battery	25/01/2021	25/01/2021
					CT983	Surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	25/01/2021	25/01/2021
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	25/01/2021	25/01/2021
							F12	Serious Injury/ Illness/ Impairment	A070504	Premature Discharge of Battery	25/01/2021	25/01/2021
							F19	Surgical Intervention	A070504	Premature Discharge of Battery	25/01/2021	25/01/2021
10393	2020	All other reportable incidents	35852	Single-chamber implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E2104	Electric Shock	A071302	Inappropriate/Inadequate Shock/Stimulation		
							F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A071302	Inappropriate/Inadequate Shock/Stimulation		
							F26	No Health Consequences or Impact	A071302	Inappropriate/Inadequate Shock/Stimulation		
10394	2020	All other reportable incidents	47265	Dual-chamber implantable pacemaker, rate-responsive	CT141	Electrodes/Leads and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem	30/12/2020	30/12/2020
									A070504	Premature Discharge of Battery	30/12/2020	30/12/2020
							F19	Surgical Intervention	A0705	Battery Problem	30/12/2020	30/12/2020
									A070504	Premature Discharge of Battery	30/12/2020	30/12/2020
							F1905	Device Revision or Replacement	A0705	Battery Problem	30/12/2020	30/12/2020
									A070504	Premature Discharge of Battery	30/12/2020	30/12/2020
					CT145	Cardiovascular devices	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem	30/12/2020	30/12/2020
									A070504	Premature Discharge of Battery	30/12/2020	30/12/2020
							F19	Surgical Intervention	A0705	Battery Problem	30/12/2020	30/12/2020
									A070504	Premature Discharge of Battery	30/12/2020	30/12/2020
							F1905	Device Revision or Replacement	A0705	Battery Problem	30/12/2020	30/12/2020
									A070504	Premature Discharge of Battery	30/12/2020	30/12/2020
					CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem	30/12/2020	30/12/2020
									A070504	Premature Discharge of Battery	30/12/2020	30/12/2020
							F19	Surgical Intervention	A0705	Battery Problem	30/12/2020	30/12/2020
									A070504	Premature Discharge of Battery	30/12/2020	30/12/2020
							F1905	Device Revision or Replacement	A0705	Battery Problem	30/12/2020	30/12/2020
									A070504	Premature Discharge of Battery	30/12/2020	30/12/2020
					CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem	30/12/2020	30/12/2020
									A070504	Premature Discharge of Battery	30/12/2020	30/12/2020
							F19	Surgical Intervention	A0705	Battery Problem	30/12/2020	30/12/2020
									A070504	Premature Discharge of Battery	30/12/2020	30/12/2020
							F1905	Device Revision or Replacement	A0705	Battery Problem	30/12/2020	30/12/2020
									A070504	Premature Discharge of Battery	30/12/2020	30/12/2020

					CT233	Surgical	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem	30/12/2020	30/12/2020
									A070504	Premature Discharge of Battery	30/12/2020	30/12/2020
							F19	Surgical Intervention	A0705	Battery Problem	30/12/2020	30/12/2020
									A070504	Premature Discharge of Battery	30/12/2020	30/12/2020
							F1905	Device Revision or Replacement	A0705	Battery Problem	30/12/2020	30/12/2020
									A070504	Premature Discharge of Battery	30/12/2020	30/12/2020
					CT2406	Implantable	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem	30/12/2020	30/12/2020
									A070504	Premature Discharge of Battery	30/12/2020	30/12/2020
							F19	Surgical Intervention	A0705	Battery Problem	30/12/2020	30/12/2020
									A070504	Premature Discharge of Battery	30/12/2020	30/12/2020
							F1905	Device Revision or Replacement	A0705	Battery Problem	30/12/2020	30/12/2020
									A070504	Premature Discharge of Battery	30/12/2020	30/12/2020
					CT2407	Active implantable	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem	30/12/2020	30/12/2020
									A070504	Premature Discharge of Battery	30/12/2020	30/12/2020
							F19	Surgical Intervention	A0705	Battery Problem	30/12/2020	30/12/2020
									A070504	Premature Discharge of Battery	30/12/2020	30/12/2020
							F1905	Device Revision or Replacement	A0705	Battery Problem	30/12/2020	30/12/2020
									A070504	Premature Discharge of Battery	30/12/2020	30/12/2020
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem	30/12/2020	30/12/2020
									A070504	Premature Discharge of Battery	30/12/2020	30/12/2020
							F19	Surgical Intervention	A0705	Battery Problem	30/12/2020	30/12/2020
									A070504	Premature Discharge of Battery	30/12/2020	30/12/2020
							F1905	Device Revision or Replacement	A0705	Battery Problem	30/12/2020	30/12/2020
									A070504	Premature Discharge of Battery	30/12/2020	30/12/2020
					CT2727	Active implant devices/ancillaries	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem	30/12/2020	30/12/2020
									A070504	Premature Discharge of Battery	30/12/2020	30/12/2020
							F19	Surgical Intervention	A0705	Battery Problem	30/12/2020	30/12/2020
									A070504	Premature Discharge of Battery	30/12/2020	30/12/2020
							F1905	Device Revision or Replacement	A0705	Battery Problem	30/12/2020	30/12/2020
									A070504	Premature Discharge of Battery	30/12/2020	30/12/2020
					CT2750	Implantable cardiac electrical stimulators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem	30/12/2020	30/12/2020
									A070504	Premature Discharge of Battery	30/12/2020	30/12/2020
							F19	Surgical Intervention	A0705	Battery Problem	30/12/2020	30/12/2020
									A070504	Premature Discharge of Battery	30/12/2020	30/12/2020
							F1905	Device Revision or Replacement	A0705	Battery Problem	30/12/2020	30/12/2020
									A070504	Premature Discharge of Battery	30/12/2020	30/12/2020
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem	30/12/2020	30/12/2020
									A070504	Premature Discharge of Battery	30/12/2020	30/12/2020
							F19	Surgical Intervention	A0705	Battery Problem	30/12/2020	30/12/2020
									A070504	Premature Discharge of Battery	30/12/2020	30/12/2020
							F1905	Device Revision or Replacement	A0705	Battery Problem	30/12/2020	30/12/2020
									A070504	Premature Discharge of Battery	30/12/2020	30/12/2020
					CT2985	Active	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem	30/12/2020	30/12/2020
									A070504	Premature Discharge of Battery	30/12/2020	30/12/2020
							F19	Surgical Intervention	A0705	Battery Problem	30/12/2020	30/12/2020
									A070504	Premature Discharge of Battery	30/12/2020	30/12/2020
							F1905	Device Revision or Replacement	A0705	Battery Problem	30/12/2020	30/12/2020
									A070504	Premature Discharge of Battery	30/12/2020	30/12/2020
					CT3001	Implantable metal	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem	30/12/2020	30/12/2020
									A070504	Premature Discharge of Battery	30/12/2020	30/12/2020
							F19	Surgical Intervention	A0705	Battery Problem	30/12/2020	30/12/2020
									A070504	Premature Discharge of Battery	30/12/2020	30/12/2020
							F1905	Device Revision or Replacement	A0705	Battery Problem	30/12/2020	30/12/2020
									A070504	Premature Discharge of Battery	30/12/2020	30/12/2020
					CT3002	Implantable synthetic polymer	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem	30/12/2020	30/12/2020
									A070504	Premature Discharge of Battery	30/12/2020	30/12/2020
							F19	Surgical Intervention	A0705	Battery Problem	30/12/2020	30/12/2020
									A070504	Premature Discharge of Battery	30/12/2020	30/12/2020
							F1905	Device Revision or Replacement	A0705	Battery Problem	30/12/2020	30/12/2020
									A070504	Premature Discharge of Battery	30/12/2020	30/12/2020
					CT3006	Implantable inorganic material	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem	30/12/2020	30/12/2020
									A070504	Premature Discharge of Battery	30/12/2020	30/12/2020
							F19	Surgical Intervention	A0705	Battery Problem	30/12/2020	30/12/2020
									A070504	Premature Discharge of Battery	30/12/2020	30/12/2020
							F1905	Device Revision or Replacement	A0705	Battery Problem	30/12/2020	30/12/2020
									A070504	Premature Discharge of Battery	30/12/2020	30/12/2020
					CT314	Electrically powered	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem	30/12/2020	30/12/2020
									A070504	Premature Discharge of Battery	30/12/2020	30/12/2020
							F19	Surgical Intervention	A0705	Battery Problem	30/12/2020	30/12/2020
									A070504	Premature Discharge of Battery	30/12/2020	30/12/2020
							F1905	Device Revision or Replacement	A0705	Battery Problem	30/12/2020	30/12/2020
									A070504	Premature Discharge of Battery	30/12/2020	30/12/2020
					CT321	Long-term surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem	30/12/2020	30/12/2020
									A070504	Premature Discharge of Battery	30/12/2020	30/12/2020
							F19	Surgical Intervention	A0705	Battery Problem	30/12/2020	30/12/2020
									A070504	Premature Discharge of Battery	30/12/2020	30/12/2020
							F1905	Device Revision or Replacement	A0705	Battery Problem	30/12/2020	30/12/2020
									A070504	Premature Discharge of Battery	30/12/2020	30/12/2020
					CT449	Generators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem	30/12/2020	30/12/2020
									A070504	Premature Discharge of Battery	30/12/2020	30/12/2020
							F19	Surgical Intervention	A0705	Battery Problem	30/12/2020	30/12/2020
									A070504	Premature Discharge of Battery	30/12/2020	30/12/2020
							F1905	Device Revision or Replacement	A0705	Battery Problem	30/12/2020	30/12/2020
									A070504	Premature Discharge of Battery	30/12/2020	30/12/2020
					CT981	Single use	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem	30/12/2020	30/12/2020
									A070504	Premature Discharge of Battery	30/12/2020	30/12/2020
							F19	Surgical Intervention	A0705	Battery Problem	30/12/2020	30/12/2020
									A070504	Premature Discharge of Battery	30/12/2020	30/12/2020

							F1905	Device Revision or Replacement	A0705	Battery Problem	30/12/2020	30/12/2020
									A070504	Premature Discharge of Battery	30/12/2020	30/12/2020
					CT983	Surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem	30/12/2020	30/12/2020
									A070504	Premature Discharge of Battery	30/12/2020	30/12/2020
							F19	Surgical Intervention	A0705	Battery Problem	30/12/2020	30/12/2020
									A070504	Premature Discharge of Battery	30/12/2020	30/12/2020
							F1905	Device Revision or Replacement	A0705	Battery Problem	30/12/2020	30/12/2020
									A070504	Premature Discharge of Battery	30/12/2020	30/12/2020
10395	2020	All other reportable incidents	35223	Endocardial/interventricular septal pacing lead	CT1684	Cardiac pacemakers and associated devices	E2103	Device Overstimulation of Tissue	A07	Electrical /Electronic Property Problem		
									A0722	Impedance Problem		
									A072201	High Impedance		
							E2403	No Clinical Signs, Symptoms or Conditions	A07	Electrical /Electronic Property Problem		
									A0722	Impedance Problem		
									A072201	High Impedance		
							F11	Non-Serious Injury/ Illness/ Impairment	A07	Electrical /Electronic Property Problem		
									A0722	Impedance Problem		
									A072201	High Impedance		
10396	2020	All other reportable incidents	35223	Endocardial/interventricular septal pacing lead	CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A0401	Break		
									A070901	Decreased Sensitivity		
									A070909	Oversensing		
									A0722	Impedance Problem		
									A072202	Low Impedance		
									A090801	Signal Artifact/Noise		
							F26	No Health Consequences or Impact	A0401	Break		
									A070901	Decreased Sensitivity		
									A070909	Oversensing		
									A0722	Impedance Problem		
									A072202	Low Impedance		
									A090801	Signal Artifact/Noise		
10397	2020	All other reportable incidents	35852	Single-chamber implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A070909	Oversensing		
									A071302	Inappropriate/Inadequate Shock/Stimulation		
									A090801	Signal Artifact/Noise		
							F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A070909	Oversensing		
									A071302	Inappropriate/Inadequate Shock/Stimulation		
									A090801	Signal Artifact/Noise		
							F26	No Health Consequences or Impact	A070909	Oversensing		
									A071302	Inappropriate/Inadequate Shock/Stimulation		
									A090801	Signal Artifact/Noise		
10398	2020	All other reportable incidents	35852	Single-chamber implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E2104	Electric Shock	A071302	Inappropriate/Inadequate Shock/Stimulation		
							F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A071302	Inappropriate/Inadequate Shock/Stimulation		
							F26	No Health Consequences or Impact	A071302	Inappropriate/Inadequate Shock/Stimulation		
10399	2020	All other reportable incidents	36241	Pacemaker/defibrillator lead	CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A051201	Device Dislodged or Dislocated		
									A070101	Failure to Capture		
									A070909	Oversensing		
							F08	Hospitalization or Prolonged Hospitalization	A051201	Device Dislodged or Dislocated		
									A070101	Failure to Capture		
									A070909	Oversensing		
							F12	Serious Injury/ Illness/ Impairment	A051201	Device Dislodged or Dislocated		
									A070101	Failure to Capture		
									A070909	Oversensing		
							F1905	Device Revision or Replacement	A051201	Device Dislodged or Dislocated		
									A070101	Failure to Capture		
									A070909	Oversensing		
10400	2020	All other reportable incidents	36241	Pacemaker/defibrillator lead	CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A0722	Impedance Problem		
									A072201	High Impedance		
							F26	No Health Consequences or Impact	A0722	Impedance Problem		
									A072201	High Impedance		
10401	2020	All other reportable incidents	35852	Single-chamber implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A070909	Oversensing		
							F26	No Health Consequences or Impact	A070909	Oversensing		
10402	2020	All other reportable incidents	37265	Dual-chamber implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A072201	High Impedance		
							F26	No Health Consequences or Impact	A072201	High Impedance		
10403	2020	All other reportable incidents	35852	Single-chamber implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A12	Connection Problem	23/11/2020	23/11/2020
							F26	No Health Consequences or Impact	A12	Connection Problem	23/11/2020	23/11/2020
10404	2020	All other reportable incidents	36241	Pacemaker/defibrillator lead	CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A05	Mechanical Problem	12/08/2020	12/08/2020
							F1905	Device Revision or Replacement	A05	Mechanical Problem	12/08/2020	12/08/2020
							F1908	Prolonged Surgery	A05	Mechanical Problem	12/08/2020	12/08/2020
10405	2020	All other reportable incidents	47270	Cardiac resynchronization therapy implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A0712	Pacing Problem		
									A072201	High Impedance		
							F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A0712	Pacing Problem		
									A072201	High Impedance		
							F26	No Health Consequences or Impact	A0712	Pacing Problem		
									A072201	High Impedance		
10406	2020	All other reportable incidents	47267	Single-chamber implantable pacemaker, rate-responsive	CT141	Electrodes/Leads and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070909	Oversensing		
									A0712	Pacing Problem		
							F26	No Health Consequences or Impact	A070909	Oversensing		
									A0712	Pacing Problem		
					CT145	Cardiovascular devices	E2403	No Clinical Signs, Symptoms or Conditions	A070909	Oversensing		
									A0712	Pacing Problem		
							F26	No Health Consequences or Impact	A070909	Oversensing		
									A0712	Pacing Problem		
					CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070909	Oversensing		
									A0712	Pacing Problem		
							F26	No Health Consequences or Impact	A070909	Oversensing		
									A0712	Pacing Problem		
					CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A070909	Oversensing		
									A0712	Pacing Problem		
							F26	No Health Consequences or Impact	A070909	Oversensing		
									A0712	Pacing Problem		

					CT233	Surgical	E2403	No Clinical Signs, Symptoms or Conditions	A070909	Oversensing		
									A0712	Pacing Problem		
							F26	No Health Consequences or Impact	A070909	Oversensing		
									A0712	Pacing Problem		
					CT2406	Implantable	E2403	No Clinical Signs, Symptoms or Conditions	A070909	Oversensing		
									A0712	Pacing Problem		
							F26	No Health Consequences or Impact	A070909	Oversensing		
									A0712	Pacing Problem		
					CT2407	Active implantable	E2403	No Clinical Signs, Symptoms or Conditions	A070909	Oversensing		
									A0712	Pacing Problem		
							F26	No Health Consequences or Impact	A070909	Oversensing		
									A0712	Pacing Problem		
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A070909	Oversensing		
									A0712	Pacing Problem		
							F26	No Health Consequences or Impact	A070909	Oversensing		
									A0712	Pacing Problem		
					CT2727	Active implant devices/ancillaries	E2403	No Clinical Signs, Symptoms or Conditions	A070909	Oversensing		
									A0712	Pacing Problem		
							F26	No Health Consequences or Impact	A070909	Oversensing		
									A0712	Pacing Problem		
					CT2750	Implantable cardiac electrical stimulators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070909	Oversensing		
									A0712	Pacing Problem		
							F26	No Health Consequences or Impact	A070909	Oversensing		
									A0712	Pacing Problem		
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A070909	Oversensing		
									A0712	Pacing Problem		
							F26	No Health Consequences or Impact	A070909	Oversensing		
									A0712	Pacing Problem		
					CT2985	Active	E2403	No Clinical Signs, Symptoms or Conditions	A070909	Oversensing		
									A0712	Pacing Problem		
							F26	No Health Consequences or Impact	A070909	Oversensing		
									A0712	Pacing Problem		
					CT3001	Implantable metal	E2403	No Clinical Signs, Symptoms or Conditions	A070909	Oversensing		
									A0712	Pacing Problem		
							F26	No Health Consequences or Impact	A070909	Oversensing		
									A0712	Pacing Problem		
					CT3002	Implantable synthetic polymer	E2403	No Clinical Signs, Symptoms or Conditions	A070909	Oversensing		
									A0712	Pacing Problem		
							F26	No Health Consequences or Impact	A070909	Oversensing		
									A0712	Pacing Problem		
					CT3006	Implantable inorganic material	E2403	No Clinical Signs, Symptoms or Conditions	A070909	Oversensing		
									A0712	Pacing Problem		
							F26	No Health Consequences or Impact	A070909	Oversensing		
									A0712	Pacing Problem		
					CT314	Electrically powered	E2403	No Clinical Signs, Symptoms or Conditions	A070909	Oversensing		
									A0712	Pacing Problem		
							F26	No Health Consequences or Impact	A070909	Oversensing		
									A0712	Pacing Problem		
					CT321	Long-term surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A070909	Oversensing		
									A0712	Pacing Problem		
							F26	No Health Consequences or Impact	A070909	Oversensing		
									A0712	Pacing Problem		
					CT449	Generators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070909	Oversensing		
									A0712	Pacing Problem		
							F26	No Health Consequences or Impact	A070909	Oversensing		
									A0712	Pacing Problem		
					CT981	Single use	E2403	No Clinical Signs, Symptoms or Conditions	A070909	Oversensing		
									A0712	Pacing Problem		
							F26	No Health Consequences or Impact	A070909	Oversensing		
									A0712	Pacing Problem		
					CT983	Surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A070909	Oversensing		
									A0712	Pacing Problem		
							F26	No Health Consequences or Impact	A070909	Oversensing		
									A0712	Pacing Problem		
10407	2020	All other reportable incidents	36241	Pacemaker/defibrillator lead	CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A072201	High Impedance		
							F26	No Health Consequences or Impact	A072201	High Impedance		
10408	2020	All other reportable incidents	37265	Dual-chamber implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	13/01/2021	13/01/2021
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	13/01/2021	13/01/2021
							F12	Serious Injury/ Illness/ Impairment	A070504	Premature Discharge of Battery	13/01/2021	13/01/2021
							F19	Surgical Intervention	A070504	Premature Discharge of Battery	13/01/2021	13/01/2021
10409	2020	All other reportable incidents	47267	Single-chamber implantable pacemaker, rate-responsive	CT141	Electrodes/Leads and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT145	Cardiovascular devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT233	Surgical	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT2406	Implantable	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT2407	Active implantable	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT2727	Active implant devices/ancillaries	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		

					CT2750	Implantable cardiac electrical stimulators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT2985	Active	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT3001	Implantable metal	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT3002	Implantable synthetic polymer	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT3006	Implantable inorganic material	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT314	Electrically powered	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT321	Long-term surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT449	Generators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT981	Single use	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT983	Surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
10410	2020	All other reportable incidents	47265	Dual-chamber implantable pacemaker, rate-responsive	CT141	Electrodes/Leads and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT145	Cardiovascular devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT233	Surgical	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT2406	Implantable	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT2407	Active implantable	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT2727	Active implant devices/ancillaries	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT2750	Implantable cardiac electrical stimulators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT2985	Active	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT3001	Implantable metal	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT3002	Implantable synthetic polymer	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT3006	Implantable inorganic material	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT314	Electrically powered	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT321	Long-term surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT449	Generators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT981	Single use	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT983	Surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
10411	2020	All other reportable incidents	36241	Pacemaker/defibrillator lead	CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A040101	Fracture		
									A070909	Oversensing		
									A072201	High Impedance		
							F26	No Health Consequences or Impact	A040101	Fracture		
									A070909	Oversensing		
									A072201	High Impedance		
10412	2020	All other reportable incidents	35852	Single-chamber implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A070909	Oversensing		
							F26	No Health Consequences or Impact	A070909	Oversensing		
10413	2020	All other reportable incidents	37265	Dual-chamber implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A0722	Impedance Problem		
							F2203	Imaging Required	A0722	Impedance Problem		
10414	2020	All other reportable incidents	36241	Pacemaker/defibrillator lead	CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A072201	High Impedance		
							F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A072201	High Impedance		
							F26	No Health Consequences or Impact	A072201	High Impedance		
10415	2020	All other reportable incidents	37265	Dual-chamber implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A0909	Unexpected Therapeutic Results		
							F26	No Health Consequences or Impact	A0909	Unexpected Therapeutic Results		
10416	2020	All other reportable incidents	47267	Single-chamber implantable pacemaker, rate-responsive	CT141	Electrodes/Leads and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	18/02/2021	18/02/2021
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	18/02/2021	18/02/2021
							F12	Serious Injury/ Illness/ Impairment	A070504	Premature Discharge of Battery	18/02/2021	18/02/2021
							F19	Surgical Intervention	A070504	Premature Discharge of Battery	18/02/2021	18/02/2021
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	18/02/2021	18/02/2021
					CT145	Cardiovascular devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	18/02/2021	18/02/2021
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	18/02/2021	18/02/2021
							F12	Serious Injury/ Illness/ Impairment	A070504	Premature Discharge of Battery	18/02/2021	18/02/2021
							F19	Surgical Intervention	A070504	Premature Discharge of Battery	18/02/2021	18/02/2021
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	18/02/2021	18/02/2021
					CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	18/02/2021	18/02/2021

							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	18/02/2021	18/02/2021
							F12	Serious Injury/ Illness/ Impairment	A070504	Premature Discharge of Battery	18/02/2021	18/02/2021
							F19	Surgical Intervention	A070504	Premature Discharge of Battery	18/02/2021	18/02/2021
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	18/02/2021	18/02/2021
					CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	18/02/2021	18/02/2021
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	18/02/2021	18/02/2021
							F12	Serious Injury/ Illness/ Impairment	A070504	Premature Discharge of Battery	18/02/2021	18/02/2021
							F19	Surgical Intervention	A070504	Premature Discharge of Battery	18/02/2021	18/02/2021
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	18/02/2021	18/02/2021
					CT233	Surgical	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	18/02/2021	18/02/2021
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	18/02/2021	18/02/2021
							F12	Serious Injury/ Illness/ Impairment	A070504	Premature Discharge of Battery	18/02/2021	18/02/2021
							F19	Surgical Intervention	A070504	Premature Discharge of Battery	18/02/2021	18/02/2021
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	18/02/2021	18/02/2021
					CT2406	Implantable	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	18/02/2021	18/02/2021
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	18/02/2021	18/02/2021
							F12	Serious Injury/ Illness/ Impairment	A070504	Premature Discharge of Battery	18/02/2021	18/02/2021
							F19	Surgical Intervention	A070504	Premature Discharge of Battery	18/02/2021	18/02/2021
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	18/02/2021	18/02/2021
					CT2407	Active implantable	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	18/02/2021	18/02/2021
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	18/02/2021	18/02/2021
							F12	Serious Injury/ Illness/ Impairment	A070504	Premature Discharge of Battery	18/02/2021	18/02/2021
							F19	Surgical Intervention	A070504	Premature Discharge of Battery	18/02/2021	18/02/2021
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	18/02/2021	18/02/2021
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	18/02/2021	18/02/2021
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	18/02/2021	18/02/2021
							F12	Serious Injury/ Illness/ Impairment	A070504	Premature Discharge of Battery	18/02/2021	18/02/2021
							F19	Surgical Intervention	A070504	Premature Discharge of Battery	18/02/2021	18/02/2021
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	18/02/2021	18/02/2021
					CT2727	Active implant devices/ancillaries	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	18/02/2021	18/02/2021
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	18/02/2021	18/02/2021
							F12	Serious Injury/ Illness/ Impairment	A070504	Premature Discharge of Battery	18/02/2021	18/02/2021
							F19	Surgical Intervention	A070504	Premature Discharge of Battery	18/02/2021	18/02/2021
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	18/02/2021	18/02/2021
					CT2750	Implantable cardiac electrical stimulators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	18/02/2021	18/02/2021
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	18/02/2021	18/02/2021
							F12	Serious Injury/ Illness/ Impairment	A070504	Premature Discharge of Battery	18/02/2021	18/02/2021
							F19	Surgical Intervention	A070504	Premature Discharge of Battery	18/02/2021	18/02/2021
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	18/02/2021	18/02/2021
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	18/02/2021	18/02/2021
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	18/02/2021	18/02/2021
							F12	Serious Injury/ Illness/ Impairment	A070504	Premature Discharge of Battery	18/02/2021	18/02/2021
							F19	Surgical Intervention	A070504	Premature Discharge of Battery	18/02/2021	18/02/2021
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	18/02/2021	18/02/2021
					CT2985	Active	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	18/02/2021	18/02/2021
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	18/02/2021	18/02/2021
							F12	Serious Injury/ Illness/ Impairment	A070504	Premature Discharge of Battery	18/02/2021	18/02/2021
							F19	Surgical Intervention	A070504	Premature Discharge of Battery	18/02/2021	18/02/2021
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	18/02/2021	18/02/2021
					CT3001	Implantable metal	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	18/02/2021	18/02/2021
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	18/02/2021	18/02/2021
							F12	Serious Injury/ Illness/ Impairment	A070504	Premature Discharge of Battery	18/02/2021	18/02/2021
							F19	Surgical Intervention	A070504	Premature Discharge of Battery	18/02/2021	18/02/2021
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	18/02/2021	18/02/2021
					CT3002	Implantable synthetic polymer	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	18/02/2021	18/02/2021
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	18/02/2021	18/02/2021
							F12	Serious Injury/ Illness/ Impairment	A070504	Premature Discharge of Battery	18/02/2021	18/02/2021
							F19	Surgical Intervention	A070504	Premature Discharge of Battery	18/02/2021	18/02/2021
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	18/02/2021	18/02/2021
					CT3006	Implantable inorganic material	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	18/02/2021	18/02/2021
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	18/02/2021	18/02/2021
							F12	Serious Injury/ Illness/ Impairment	A070504	Premature Discharge of Battery	18/02/2021	18/02/2021
							F19	Surgical Intervention	A070504	Premature Discharge of Battery	18/02/2021	18/02/2021
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	18/02/2021	18/02/2021
					CT314	Electrically powered	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	18/02/2021	18/02/2021
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	18/02/2021	18/02/2021
							F12	Serious Injury/ Illness/ Impairment	A070504	Premature Discharge of Battery	18/02/2021	18/02/2021
							F19	Surgical Intervention	A070504	Premature Discharge of Battery	18/02/2021	18/02/2021
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	18/02/2021	18/02/2021
					CT321	Long-term surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	18/02/2021	18/02/2021
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	18/02/2021	18/02/2021
							F12	Serious Injury/ Illness/ Impairment	A070504	Premature Discharge of Battery	18/02/2021	18/02/2021
							F19	Surgical Intervention	A070504	Premature Discharge of Battery	18/02/2021	18/02/2021
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	18/02/2021	18/02/2021
					CT449	Generators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	18/02/2021	18/02/2021
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	18/02/2021	18/02/2021
							F12	Serious Injury/ Illness/ Impairment	A070504	Premature Discharge of Battery	18/02/2021	18/02/2021
							F19	Surgical Intervention	A070504	Premature Discharge of Battery	18/02/2021	18/02/2021
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	18/02/2021	18/02/2021
					CT981	Single use	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	18/02/2021	18/02/2021
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	18/02/2021	18/02/2021
							F12	Serious Injury/ Illness/ Impairment	A070504	Premature Discharge of Battery	18/02/2021	18/02/2021
							F19	Surgical Intervention	A070504	Premature Discharge of Battery	18/02/2021	18/02/2021
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	18/02/2021	18/02/2021
					CT983	Surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	18/02/2021	18/02/2021
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	18/02/2021	18/02/2021
							F12	Serious Injury/ Illness/ Impairment	A070504	Premature Discharge of Battery	18/02/2021	18/02/2021
							F19	Surgical Intervention	A070504	Premature Discharge of Battery	18/02/2021	18/02/2021

							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	18/02/2021	18/02/2021
10417	2020	All other reportable incidents	36241	Pacemaker/defibrillator lead	CT1684	Cardiac pacemakers and associated devices	E2401	Insufficient Information	A072201	High Impedance		
							E2403	No Clinical Signs, Symptoms or Conditions	A072201	High Impedance		
							F26	No Health Consequences or Impact	A072201	High Impedance		
10418	2020	All other reportable incidents	47265	Dual-chamber implantable pacemaker, rate-responsive	CT141	Electrodes/Leads and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A0508	Noise, Audible		
									A0722	Impedance Problem		
							F26	No Health Consequences or Impact	A0508	Noise, Audible		
									A0722	Impedance Problem		
					CT145	Cardiovascular devices	E2403	No Clinical Signs, Symptoms or Conditions	A0508	Noise, Audible		
									A0722	Impedance Problem		
							F26	No Health Consequences or Impact	A0508	Noise, Audible		
									A0722	Impedance Problem		
					CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A0508	Noise, Audible		
									A0722	Impedance Problem		
							F26	No Health Consequences or Impact	A0508	Noise, Audible		
									A0722	Impedance Problem		
					CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A0508	Noise, Audible		
									A0722	Impedance Problem		
							F26	No Health Consequences or Impact	A0508	Noise, Audible		
									A0722	Impedance Problem		
					CT233	Surgical	E2403	No Clinical Signs, Symptoms or Conditions	A0508	Noise, Audible		
									A0722	Impedance Problem		
							F26	No Health Consequences or Impact	A0508	Noise, Audible		
									A0722	Impedance Problem		
					CT2406	Implantable	E2403	No Clinical Signs, Symptoms or Conditions	A0508	Noise, Audible		
									A0722	Impedance Problem		
							F26	No Health Consequences or Impact	A0508	Noise, Audible		
									A0722	Impedance Problem		
					CT2407	Active implantable	E2403	No Clinical Signs, Symptoms or Conditions	A0508	Noise, Audible		
									A0722	Impedance Problem		
							F26	No Health Consequences or Impact	A0508	Noise, Audible		
									A0722	Impedance Problem		
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A0508	Noise, Audible		
									A0722	Impedance Problem		
							F26	No Health Consequences or Impact	A0508	Noise, Audible		
									A0722	Impedance Problem		
					CT2727	Active implant devices/ancillaries	E2403	No Clinical Signs, Symptoms or Conditions	A0508	Noise, Audible		
									A0722	Impedance Problem		
							F26	No Health Consequences or Impact	A0508	Noise, Audible		
									A0722	Impedance Problem		
					CT2750	Implantable cardiac electrical stimulators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A0508	Noise, Audible		
									A0722	Impedance Problem		
							F26	No Health Consequences or Impact	A0508	Noise, Audible		
									A0722	Impedance Problem		
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A0508	Noise, Audible		
									A0722	Impedance Problem		
							F26	No Health Consequences or Impact	A0508	Noise, Audible		
									A0722	Impedance Problem		
					CT2985	Active	E2403	No Clinical Signs, Symptoms or Conditions	A0508	Noise, Audible		
									A0722	Impedance Problem		
							F26	No Health Consequences or Impact	A0508	Noise, Audible		
									A0722	Impedance Problem		
					CT3001	Implantable metal	E2403	No Clinical Signs, Symptoms or Conditions	A0508	Noise, Audible		
									A0722	Impedance Problem		
							F26	No Health Consequences or Impact	A0508	Noise, Audible		
									A0722	Impedance Problem		
					CT3002	Implantable synthetic polymer	E2403	No Clinical Signs, Symptoms or Conditions	A0508	Noise, Audible		
									A0722	Impedance Problem		
							F26	No Health Consequences or Impact	A0508	Noise, Audible		
									A0722	Impedance Problem		
					CT3006	Implantable inorganic material	E2403	No Clinical Signs, Symptoms or Conditions	A0508	Noise, Audible		
									A0722	Impedance Problem		
							F26	No Health Consequences or Impact	A0508	Noise, Audible		
									A0722	Impedance Problem		
					CT314	Electrically powered	E2403	No Clinical Signs, Symptoms or Conditions	A0508	Noise, Audible		
									A0722	Impedance Problem		
							F26	No Health Consequences or Impact	A0508	Noise, Audible		
									A0722	Impedance Problem		
					CT321	Long-term surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A0508	Noise, Audible		
									A0722	Impedance Problem		
							F26	No Health Consequences or Impact	A0508	Noise, Audible		
									A0722	Impedance Problem		
					CT449	Generators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A0508	Noise, Audible		
									A0722	Impedance Problem		
							F26	No Health Consequences or Impact	A0508	Noise, Audible		
									A0722	Impedance Problem		
					CT981	Single use	E2403	No Clinical Signs, Symptoms or Conditions	A0508	Noise, Audible		
									A0722	Impedance Problem		
							F26	No Health Consequences or Impact	A0508	Noise, Audible		
									A0722	Impedance Problem		
					CT983	Surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A0508	Noise, Audible		
									A0722	Impedance Problem		
							F26	No Health Consequences or Impact	A0508	Noise, Audible		
									A0722	Impedance Problem		
10419	2020	All other reportable incidents	47265	Dual-chamber implantable pacemaker, rate-responsive	CT141	Electrodes/Leads and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070101	Failure to Capture		
									A070504	Premature Discharge of Battery		
									A072201	High Impedance		
									A090801	Signal Artifact/Noise		

						F26	No Health Consequences or Impact	A070101	Failure to Capture		
								A070504	Premature Discharge of Battery		
								A072201	High Impedance		
								A090801	Signal Artifact/Noise		
					CT145	Cardiovascular devices	E2403	No Clinical Signs, Symptoms or Conditions	A070101	Failure to Capture		
									A070504	Premature Discharge of Battery		
									A072201	High Impedance		
									A090801	Signal Artifact/Noise		
							F26	No Health Consequences or Impact	A070101	Failure to Capture		
									A070504	Premature Discharge of Battery		
									A072201	High Impedance		
									A090801	Signal Artifact/Noise		
					CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070101	Failure to Capture		
									A070504	Premature Discharge of Battery		
									A072201	High Impedance		
									A090801	Signal Artifact/Noise		
							F26	No Health Consequences or Impact	A070101	Failure to Capture		
									A070504	Premature Discharge of Battery		
									A072201	High Impedance		
									A090801	Signal Artifact/Noise		
					CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A070101	Failure to Capture		
									A070504	Premature Discharge of Battery		
									A072201	High Impedance		
									A090801	Signal Artifact/Noise		
							F26	No Health Consequences or Impact	A070101	Failure to Capture		
									A070504	Premature Discharge of Battery		
									A072201	High Impedance		
									A090801	Signal Artifact/Noise		
					CT233	Surgical	E2403	No Clinical Signs, Symptoms or Conditions	A070101	Failure to Capture		
									A070504	Premature Discharge of Battery		
									A072201	High Impedance		
									A090801	Signal Artifact/Noise		
							F26	No Health Consequences or Impact	A070101	Failure to Capture		
									A070504	Premature Discharge of Battery		
									A072201	High Impedance		
									A090801	Signal Artifact/Noise		
					CT2406	Implantable	E2403	No Clinical Signs, Symptoms or Conditions	A070101	Failure to Capture		
									A070504	Premature Discharge of Battery		
									A072201	High Impedance		
									A090801	Signal Artifact/Noise		
							F26	No Health Consequences or Impact	A070101	Failure to Capture		
									A070504	Premature Discharge of Battery		
									A072201	High Impedance		
									A090801	Signal Artifact/Noise		
					CT2407	Active implantable	E2403	No Clinical Signs, Symptoms or Conditions	A070101	Failure to Capture		
									A070504	Premature Discharge of Battery		
									A072201	High Impedance		
									A090801	Signal Artifact/Noise		
							F26	No Health Consequences or Impact	A070101	Failure to Capture		
									A070504	Premature Discharge of Battery		
									A072201	High Impedance		
									A090801	Signal Artifact/Noise		
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A070101	Failure to Capture		
									A070504	Premature Discharge of Battery		
									A072201	High Impedance		
									A090801	Signal Artifact/Noise		
							F26	No Health Consequences or Impact	A070101	Failure to Capture		
									A070504	Premature Discharge of Battery		
									A072201	High Impedance		
									A090801	Signal Artifact/Noise		
					CT2727	Active implant devices/ancillaries	E2403	No Clinical Signs, Symptoms or Conditions	A070101	Failure to Capture		
									A070504	Premature Discharge of Battery		
									A072201	High Impedance		
									A090801	Signal Artifact/Noise		
							F26	No Health Consequences or Impact	A070101	Failure to Capture		
									A070504	Premature Discharge of Battery		
									A072201	High Impedance		
									A090801	Signal Artifact/Noise		
					CT2750	Implantable cardiac electrical stimulators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070101	Failure to Capture		
									A070504	Premature Discharge of Battery		
									A072201	High Impedance		
									A090801	Signal Artifact/Noise		
							F26	No Health Consequences or Impact	A070101	Failure to Capture		
									A070504	Premature Discharge of Battery		
									A072201	High Impedance		
									A090801	Signal Artifact/Noise		
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A070101	Failure to Capture		
									A070504	Premature Discharge of Battery		
									A072201	High Impedance		
									A090801	Signal Artifact/Noise		
							F26	No Health Consequences or Impact	A070101	Failure to Capture		
									A070504	Premature Discharge of Battery		
									A072201	High Impedance		
									A090801	Signal Artifact/Noise		
					CT2985	Active	E2403	No Clinical Signs, Symptoms or Conditions	A070101	Failure to Capture		
									A070504	Premature Discharge of Battery		
									A072201	High Impedance		
									A090801	Signal Artifact/Noise		

							F26	No Health Consequences or Impact	A070101	Failure to Capture			
									A070504	Premature Discharge of Battery			
									A072201	High Impedance			
									A090801	Signal Artifact/Noise			
					CT3001	Implantable metal	E2403	No Clinical Signs, Symptoms or Conditions	A070101	Failure to Capture			
									A070504	Premature Discharge of Battery			
									A072201	High Impedance			
									A090801	Signal Artifact/Noise			
							F26	No Health Consequences or Impact	A070101	Failure to Capture			
									A070504	Premature Discharge of Battery			
									A072201	High Impedance			
									A090801	Signal Artifact/Noise			
					CT3002	Implantable synthetic polymer	E2403	No Clinical Signs, Symptoms or Conditions	A070101	Failure to Capture			
									A070504	Premature Discharge of Battery			
									A072201	High Impedance			
									A090801	Signal Artifact/Noise			
							F26	No Health Consequences or Impact	A070101	Failure to Capture			
									A070504	Premature Discharge of Battery			
									A072201	High Impedance			
									A090801	Signal Artifact/Noise			
					CT3006	Implantable inorganic material	E2403	No Clinical Signs, Symptoms or Conditions	A070101	Failure to Capture			
									A070504	Premature Discharge of Battery			
									A072201	High Impedance			
									A090801	Signal Artifact/Noise			
							F26	No Health Consequences or Impact	A070101	Failure to Capture			
									A070504	Premature Discharge of Battery			
									A072201	High Impedance			
									A090801	Signal Artifact/Noise			
					CT314	Electrically powered	E2403	No Clinical Signs, Symptoms or Conditions	A070101	Failure to Capture			
									A070504	Premature Discharge of Battery			
									A072201	High Impedance			
									A090801	Signal Artifact/Noise			
							F26	No Health Consequences or Impact	A070101	Failure to Capture			
									A070504	Premature Discharge of Battery			
									A072201	High Impedance			
									A090801	Signal Artifact/Noise			
					CT321	Long-term surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A070101	Failure to Capture			
									A070504	Premature Discharge of Battery			
									A072201	High Impedance			
									A090801	Signal Artifact/Noise			
							F26	No Health Consequences or Impact	A070101	Failure to Capture			
									A070504	Premature Discharge of Battery			
									A072201	High Impedance			
									A090801	Signal Artifact/Noise			
					CT449	Generators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070101	Failure to Capture			
									A070504	Premature Discharge of Battery			
									A072201	High Impedance			
									A090801	Signal Artifact/Noise			
							F26	No Health Consequences or Impact	A070101	Failure to Capture			
									A070504	Premature Discharge of Battery			
									A072201	High Impedance			
									A090801	Signal Artifact/Noise			
					CT981	Single use	E2403	No Clinical Signs, Symptoms or Conditions	A070101	Failure to Capture			
									A070504	Premature Discharge of Battery			
									A072201	High Impedance			
									A090801	Signal Artifact/Noise			
							F26	No Health Consequences or Impact	A070101	Failure to Capture			
									A070504	Premature Discharge of Battery			
									A072201	High Impedance			
									A090801	Signal Artifact/Noise			
					CT983	Surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A070101	Failure to Capture			
									A070504	Premature Discharge of Battery			
									A072201	High Impedance			
									A090801	Signal Artifact/Noise			
							F26	No Health Consequences or Impact	A070101	Failure to Capture			
									A070504	Premature Discharge of Battery			
									A072201	High Impedance			
									A090801	Signal Artifact/Noise			
10420	2020	All other reportable incidents	36241	Pacemaker/defibrillator lead	CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A072201	High Impedance			
							F24	Insufficient Health Impact Information	A072201	High Impedance			
10421	2020	All other reportable incidents	35223	Endocardial/interventricular septal pacing lead	CT1684	Cardiac pacemakers and associated devices	E0138	Undesired Nerve Stimulation	A24	No Apparent Device Problem	18/09/2017	18/09/2017	
							E0604	Cardiac Perforation	A24	No Apparent Device Problem	18/09/2017	18/09/2017	
							E061202	Myocardial Infarction	A24	No Apparent Device Problem	18/09/2017	18/09/2017	
							F1203	Life Threatening Illness or Injury	A24	No Apparent Device Problem	18/09/2017	18/09/2017	
							F1905	Device Revision or Replacement	A24	No Apparent Device Problem	18/09/2017	18/09/2017	
10422	2020	All other reportable incidents	35852	Single-chamber implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	17/02/2021	17/02/2021	
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	17/02/2021	17/02/2021	
							F19	Surgical Intervention	A070504	Premature Discharge of Battery	17/02/2021	17/02/2021	
10423	2020	All other reportable incidents	35852	Single-chamber implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E2104	Electric Shock	A071302	Inappropriate/Inadequate Shock/Stimulation			
							F08	Hospitalization or Prolonged Hospitalization	A071302	Inappropriate/Inadequate Shock/Stimulation			
							F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A071302	Inappropriate/Inadequate Shock/Stimulation			
10424	2020	All other reportable incidents	36241	Pacemaker/defibrillator lead	CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A051201	Device Dislodged or Dislocated			
									A070101	Failure to Capture			
							F15	Recognized Device or Procedural Complication	A051201	Device Dislodged or Dislocated			
									A070101	Failure to Capture			
10425	2020	All other reportable incidents	47270	Cardiac resynchronization therapy implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	14/12/2020	14/12/2020	
							F1901	Additional Surgery	A070504	Premature Discharge of Battery	14/12/2020	14/12/2020	
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	14/12/2020	14/12/2020	

10426	2020	All other reportable incidents	37265	Dual-chamber implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E2104	Electric Shock	A0712	Pacing Problem		
									A071302	Inappropriate/Inadequate Shock/Stimulation		
									A090801	Signal Artifact/Noise		
							E2403	No Clinical Signs, Symptoms or Conditions	A0712	Pacing Problem		
									A071302	Inappropriate/Inadequate Shock/Stimulation		
									A090801	Signal Artifact/Noise		
							F08	Hospitalization or Prolonged Hospitalization	A0712	Pacing Problem		
									A071302	Inappropriate/Inadequate Shock/Stimulation		
									A090801	Signal Artifact/Noise		
							F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A0712	Pacing Problem		
									A071302	Inappropriate/Inadequate Shock/Stimulation		
									A090801	Signal Artifact/Noise		
							F12	Serious Injury/ Illness/ Impairment	A0712	Pacing Problem		
									A071302	Inappropriate/Inadequate Shock/Stimulation		
									A090801	Signal Artifact/Noise		
							F15	Recognized Device or Procedural Complication	A0712	Pacing Problem		
									A071302	Inappropriate/Inadequate Shock/Stimulation		
									A090801	Signal Artifact/Noise		
10427	2020	All other reportable incidents	35852	Single-chamber implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E2401	Insufficient Information	A160601	Premature Elective Replacement Indicator		
							F1905	Device Revision or Replacement	A160601	Premature Elective Replacement Indicator		
							F24	Insufficient Health Impact Information	A160601	Premature Elective Replacement Indicator		
10428	2020	All other reportable incidents	17882	Physiologic-monitoring defibrillation system	CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070803	Failure to Power Up		
							F27	Problem Identified During Non-Clinical Procedure	A070803	Failure to Power Up		
10429	2020	All other reportable incidents	47265	Dual-chamber implantable pacemaker, rate-responsive	CT141	Electrodes/Leads and associated devices	E0601	Arrhythmia	A070504	Premature Discharge of Battery	01/07/2020	
									A071205	Pacemaker Found in Back-Up Mode	01/07/2020	
							F1903	Device Explantation	A070504	Premature Discharge of Battery	01/07/2020	
									A071205	Pacemaker Found in Back-Up Mode	01/07/2020	
							F21	Unexpected Deterioration	A070504	Premature Discharge of Battery	01/07/2020	
									A071205	Pacemaker Found in Back-Up Mode	01/07/2020	
					CT145	Cardiovascular devices	E0601	Arrhythmia	A070504	Premature Discharge of Battery	01/07/2020	
									A071205	Pacemaker Found in Back-Up Mode	01/07/2020	
							F1903	Device Explantation	A070504	Premature Discharge of Battery	01/07/2020	
									A071205	Pacemaker Found in Back-Up Mode	01/07/2020	
							F21	Unexpected Deterioration	A070504	Premature Discharge of Battery	01/07/2020	
									A071205	Pacemaker Found in Back-Up Mode	01/07/2020	
					CT1684	Cardiac pacemakers and associated devices	E0601	Arrhythmia	A070504	Premature Discharge of Battery	01/07/2020	
									A071205	Pacemaker Found in Back-Up Mode	01/07/2020	
							F1903	Device Explantation	A070504	Premature Discharge of Battery	01/07/2020	
									A071205	Pacemaker Found in Back-Up Mode	01/07/2020	
							F21	Unexpected Deterioration	A070504	Premature Discharge of Battery	01/07/2020	
									A071205	Pacemaker Found in Back-Up Mode	01/07/2020	
					CT1685	Cardiac pacemakers	E0601	Arrhythmia	A070504	Premature Discharge of Battery	01/07/2020	
									A071205	Pacemaker Found in Back-Up Mode	01/07/2020	
							F1903	Device Explantation	A070504	Premature Discharge of Battery	01/07/2020	
									A071205	Pacemaker Found in Back-Up Mode	01/07/2020	
							F21	Unexpected Deterioration	A070504	Premature Discharge of Battery	01/07/2020	
									A071205	Pacemaker Found in Back-Up Mode	01/07/2020	
					CT233	Surgical	E0601	Arrhythmia	A070504	Premature Discharge of Battery	01/07/2020	
									A071205	Pacemaker Found in Back-Up Mode	01/07/2020	
							F1903	Device Explantation	A070504	Premature Discharge of Battery	01/07/2020	
									A071205	Pacemaker Found in Back-Up Mode	01/07/2020	
							F21	Unexpected Deterioration	A070504	Premature Discharge of Battery	01/07/2020	
									A071205	Pacemaker Found in Back-Up Mode	01/07/2020	
					CT2406	Implantable	E0601	Arrhythmia	A070504	Premature Discharge of Battery	01/07/2020	
									A071205	Pacemaker Found in Back-Up Mode	01/07/2020	
							F1903	Device Explantation	A070504	Premature Discharge of Battery	01/07/2020	
									A071205	Pacemaker Found in Back-Up Mode	01/07/2020	
							F21	Unexpected Deterioration	A070504	Premature Discharge of Battery	01/07/2020	
									A071205	Pacemaker Found in Back-Up Mode	01/07/2020	
					CT2407	Active implantable	E0601	Arrhythmia	A070504	Premature Discharge of Battery	01/07/2020	
									A071205	Pacemaker Found in Back-Up Mode	01/07/2020	
							F1903	Device Explantation	A070504	Premature Discharge of Battery	01/07/2020	
									A071205	Pacemaker Found in Back-Up Mode	01/07/2020	
							F21	Unexpected Deterioration	A070504	Premature Discharge of Battery	01/07/2020	
									A071205	Pacemaker Found in Back-Up Mode	01/07/2020	
					CT257	Implantable cardiac pacemakers	E0601	Arrhythmia	A070504	Premature Discharge of Battery	01/07/2020	
									A071205	Pacemaker Found in Back-Up Mode	01/07/2020	
							F1903	Device Explantation	A070504	Premature Discharge of Battery	01/07/2020	
									A071205	Pacemaker Found in Back-Up Mode	01/07/2020	
							F21	Unexpected Deterioration	A070504	Premature Discharge of Battery	01/07/2020	
									A071205	Pacemaker Found in Back-Up Mode	01/07/2020	
					CT2727	Active implant devices/ancillaries	E0601	Arrhythmia	A070504	Premature Discharge of Battery	01/07/2020	
									A071205	Pacemaker Found in Back-Up Mode	01/07/2020	
							F1903	Device Explantation	A070504	Premature Discharge of Battery	01/07/2020	
									A071205	Pacemaker Found in Back-Up Mode	01/07/2020	
							F21	Unexpected Deterioration	A070504	Premature Discharge of Battery	01/07/2020	
									A071205	Pacemaker Found in Back-Up Mode	01/07/2020	
					CT2750	Implantable cardiac electrical stimulators and associated devices	E0601	Arrhythmia	A070504	Premature Discharge of Battery	01/07/2020	
									A071205	Pacemaker Found in Back-Up Mode	01/07/2020	
							F1903	Device Explantation	A070504	Premature Discharge of Battery	01/07/2020	
									A071205	Pacemaker Found in Back-Up Mode	01/07/2020	
							F21	Unexpected Deterioration	A070504	Premature Discharge of Battery	01/07/2020	
									A071205	Pacemaker Found in Back-Up Mode	01/07/2020	
					CT2866	Implantable cardiac electrical stimulators	E0601	Arrhythmia	A070504	Premature Discharge of Battery	01/07/2020	
									A071205	Pacemaker Found in Back-Up Mode	01/07/2020	
							F1903	Device Explantation	A070504	Premature Discharge of Battery	01/07/2020	
									A071205	Pacemaker Found in Back-Up Mode	01/07/2020	
							F21	Unexpected Deterioration	A070504	Premature Discharge of Battery	01/07/2020	

									A071205	Pacemaker Found in Back-Up Mode	01/07/2020	
					CT2985	Active	E0601	Arrhythmia	A070504	Premature Discharge of Battery	01/07/2020	
									A071205	Pacemaker Found in Back-Up Mode	01/07/2020	
							F1903	Device Explantation	A070504	Premature Discharge of Battery	01/07/2020	
									A071205	Pacemaker Found in Back-Up Mode	01/07/2020	
							F21	Unexpected Deterioration	A070504	Premature Discharge of Battery	01/07/2020	
									A071205	Pacemaker Found in Back-Up Mode	01/07/2020	
					CT3001	Implantable metal	E0601	Arrhythmia	A070504	Premature Discharge of Battery	01/07/2020	
									A071205	Pacemaker Found in Back-Up Mode	01/07/2020	
							F1903	Device Explantation	A070504	Premature Discharge of Battery	01/07/2020	
									A071205	Pacemaker Found in Back-Up Mode	01/07/2020	
							F21	Unexpected Deterioration	A070504	Premature Discharge of Battery	01/07/2020	
									A071205	Pacemaker Found in Back-Up Mode	01/07/2020	
					CT3002	Implantable synthetic polymer	E0601	Arrhythmia	A070504	Premature Discharge of Battery	01/07/2020	
									A071205	Pacemaker Found in Back-Up Mode	01/07/2020	
							F1903	Device Explantation	A070504	Premature Discharge of Battery	01/07/2020	
									A071205	Pacemaker Found in Back-Up Mode	01/07/2020	
							F21	Unexpected Deterioration	A070504	Premature Discharge of Battery	01/07/2020	
									A071205	Pacemaker Found in Back-Up Mode	01/07/2020	
					CT3006	Implantable inorganic material	E0601	Arrhythmia	A070504	Premature Discharge of Battery	01/07/2020	
									A071205	Pacemaker Found in Back-Up Mode	01/07/2020	
							F1903	Device Explantation	A070504	Premature Discharge of Battery	01/07/2020	
									A071205	Pacemaker Found in Back-Up Mode	01/07/2020	
							F21	Unexpected Deterioration	A070504	Premature Discharge of Battery	01/07/2020	
									A071205	Pacemaker Found in Back-Up Mode	01/07/2020	
					CT314	Electrically powered	E0601	Arrhythmia	A070504	Premature Discharge of Battery	01/07/2020	
									A071205	Pacemaker Found in Back-Up Mode	01/07/2020	
							F1903	Device Explantation	A070504	Premature Discharge of Battery	01/07/2020	
									A071205	Pacemaker Found in Back-Up Mode	01/07/2020	
							F21	Unexpected Deterioration	A070504	Premature Discharge of Battery	01/07/2020	
									A071205	Pacemaker Found in Back-Up Mode	01/07/2020	
					CT321	Long-term surgical invasive	E0601	Arrhythmia	A070504	Premature Discharge of Battery	01/07/2020	
									A071205	Pacemaker Found in Back-Up Mode	01/07/2020	
							F1903	Device Explantation	A070504	Premature Discharge of Battery	01/07/2020	
									A071205	Pacemaker Found in Back-Up Mode	01/07/2020	
							F21	Unexpected Deterioration	A070504	Premature Discharge of Battery	01/07/2020	
									A071205	Pacemaker Found in Back-Up Mode	01/07/2020	
					CT449	Generators and associated devices	E0601	Arrhythmia	A070504	Premature Discharge of Battery	01/07/2020	
									A071205	Pacemaker Found in Back-Up Mode	01/07/2020	
							F1903	Device Explantation	A070504	Premature Discharge of Battery	01/07/2020	
									A071205	Pacemaker Found in Back-Up Mode	01/07/2020	
							F21	Unexpected Deterioration	A070504	Premature Discharge of Battery	01/07/2020	
									A071205	Pacemaker Found in Back-Up Mode	01/07/2020	
					CT981	Single use	E0601	Arrhythmia	A070504	Premature Discharge of Battery	01/07/2020	
									A071205	Pacemaker Found in Back-Up Mode	01/07/2020	
							F1903	Device Explantation	A070504	Premature Discharge of Battery	01/07/2020	
									A071205	Pacemaker Found in Back-Up Mode	01/07/2020	
							F21	Unexpected Deterioration	A070504	Premature Discharge of Battery	01/07/2020	
									A071205	Pacemaker Found in Back-Up Mode	01/07/2020	
					CT983	Surgical invasive	E0601	Arrhythmia	A070504	Premature Discharge of Battery	01/07/2020	
									A071205	Pacemaker Found in Back-Up Mode	01/07/2020	
							F1903	Device Explantation	A070504	Premature Discharge of Battery	01/07/2020	
									A071205	Pacemaker Found in Back-Up Mode	01/07/2020	
							F21	Unexpected Deterioration	A070504	Premature Discharge of Battery	01/07/2020	
									A071205	Pacemaker Found in Back-Up Mode	01/07/2020	
10430	2020	All other reportable incidents	17882	Physiologic-monitoring defibrillation system	CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A0719	Unexpected Shutdown		
							F26	No Health Consequences or Impact	A0719	Unexpected Shutdown		
10431	2020	All other reportable incidents	63188	Left-ventricular wireless pacing system sensor/transmitter	CT141	Electrodes/Leads and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070801	Complete Loss of Power		
					CT145	Cardiovascular devices	E2403	No Clinical Signs, Symptoms or Conditions	A070801	Complete Loss of Power		
					CT1577	Wireless monitoring system transmitters	E2403	No Clinical Signs, Symptoms or Conditions	A070801	Complete Loss of Power		
					CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070801	Complete Loss of Power		
					CT2317	Transmitters/Receivers/Transceivers	E2403	No Clinical Signs, Symptoms or Conditions	A070801	Complete Loss of Power		
					CT233	Surgical	E2403	No Clinical Signs, Symptoms or Conditions	A070801	Complete Loss of Power		
					CT2406	Implantable	E2403	No Clinical Signs, Symptoms or Conditions	A070801	Complete Loss of Power		
					CT2407	Active implantable	E2403	No Clinical Signs, Symptoms or Conditions	A070801	Complete Loss of Power		
					CT255	Transducers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070801	Complete Loss of Power		
					CT2726	Physiological monitoring/measuring devices	E2403	No Clinical Signs, Symptoms or Conditions	A070801	Complete Loss of Power		
					CT2727	Active implant devices/ancillaries	E2403	No Clinical Signs, Symptoms or Conditions	A070801	Complete Loss of Power		
					CT2750	Implantable cardiac electrical stimulators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070801	Complete Loss of Power		
					CT2885	Physiological monitoring/measuring system transmitters/receivers/transceivers	E2403	No Clinical Signs, Symptoms or Conditions	A070801	Complete Loss of Power		
					CT2985	Active	E2403	No Clinical Signs, Symptoms or Conditions	A070801	Complete Loss of Power		
					CT3001	Implantable metal	E2403	No Clinical Signs, Symptoms or Conditions	A070801	Complete Loss of Power		
					CT3002	Implantable synthetic polymer	E2403	No Clinical Signs, Symptoms or Conditions	A070801	Complete Loss of Power		
					CT3006	Implantable inorganic material	E2403	No Clinical Signs, Symptoms or Conditions	A070801	Complete Loss of Power		
					CT314	Electrically powered	E2403	No Clinical Signs, Symptoms or Conditions	A070801	Complete Loss of Power		
					CT321	Long-term surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A070801	Complete Loss of Power		
					CT384	Sensors and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070801	Complete Loss of Power		
					CT449	Generators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070801	Complete Loss of Power		
					CT981	Single use	E2403	No Clinical Signs, Symptoms or Conditions	A070801	Complete Loss of Power		
					CT983	Surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A070801	Complete Loss of Power		
					CT994	Sensors	E2403	No Clinical Signs, Symptoms or Conditions	A070801	Complete Loss of Power		
10432	2020	All other reportable incidents	47263	Cardiac resynchronization therapy implantable pacemaker	CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		

							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
10433	2020	All other reportable incidents	17882	Physiologic-monitoring defibrillation system	CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A0719	Unexpected Shutdown		
							F26	No Health Consequences or Impact	A0719	Unexpected Shutdown		
10434	2020	All other reportable incidents	47270	Cardiac resynchronization therapy implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	07/12/2020	07/12/2020
							F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A070504	Premature Discharge of Battery	07/12/2020	07/12/2020
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	07/12/2020	07/12/2020
10435	2020	All other reportable incidents	37265	Dual-chamber implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A0203	Defective Device	26/02/2021	26/02/2021
									A070504	Premature Discharge of Battery	26/02/2021	26/02/2021
							F08	Hospitalization or Prolonged Hospitalization	A0203	Defective Device	26/02/2021	26/02/2021
									A070504	Premature Discharge of Battery	26/02/2021	26/02/2021
							F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A0203	Defective Device	26/02/2021	26/02/2021
									A070504	Premature Discharge of Battery	26/02/2021	26/02/2021
							F1905	Device Revision or Replacement	A0203	Defective Device	26/02/2021	26/02/2021
									A070504	Premature Discharge of Battery	26/02/2021	26/02/2021
10436	2020	All other reportable incidents	35223	Endocardial/interventricular septal pacing lead	CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A0401	Break		
									A0709	Device Sensing Problem		
									A070909	Oversensing		
									A0712	Pacing Problem		
									A090801	Signal Artifact/Noise		
							F1901	Additional Surgery	A0401	Break		
									A0709	Device Sensing Problem		
									A070909	Oversensing		
									A0712	Pacing Problem		
									A090801	Signal Artifact/Noise		
10437	2020	All other reportable incidents	63188	Left-ventricular wireless pacing system sensor/transmitter	CT141	Electrodes/Leads and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070801	Complete Loss of Power		
					CT145	Cardiovascular devices	E2403	No Clinical Signs, Symptoms or Conditions	A070801	Complete Loss of Power		
					CT1577	Wireless monitoring system transmitters	E2403	No Clinical Signs, Symptoms or Conditions	A070801	Complete Loss of Power		
					CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070801	Complete Loss of Power		
					CT2317	Transmitters/Receivers/Transceivers	E2403	No Clinical Signs, Symptoms or Conditions	A070801	Complete Loss of Power		
					CT233	Surgical	E2403	No Clinical Signs, Symptoms or Conditions	A070801	Complete Loss of Power		
					CT2406	Implantable	E2403	No Clinical Signs, Symptoms or Conditions	A070801	Complete Loss of Power		
					CT2407	Active implantable	E2403	No Clinical Signs, Symptoms or Conditions	A070801	Complete Loss of Power		
					CT255	Transducers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070801	Complete Loss of Power		
					CT2726	Physiological monitoring/measuring devices	E2403	No Clinical Signs, Symptoms or Conditions	A070801	Complete Loss of Power		
					CT2727	Active implant devices/ancillaries	E2403	No Clinical Signs, Symptoms or Conditions	A070801	Complete Loss of Power		
					CT2750	Implantable cardiac electrical stimulators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070801	Complete Loss of Power		
					CT2885	Physiological monitoring/measuring system transmitters/receivers/transceivers	E2403	No Clinical Signs, Symptoms or Conditions	A070801	Complete Loss of Power		
					CT2985	Active	E2403	No Clinical Signs, Symptoms or Conditions	A070801	Complete Loss of Power		
					CT3001	Implantable metal	E2403	No Clinical Signs, Symptoms or Conditions	A070801	Complete Loss of Power		
					CT3002	Implantable synthetic polymer	E2403	No Clinical Signs, Symptoms or Conditions	A070801	Complete Loss of Power		
					CT3006	Implantable inorganic material	E2403	No Clinical Signs, Symptoms or Conditions	A070801	Complete Loss of Power		
					CT314	Electrically powered	E2403	No Clinical Signs, Symptoms or Conditions	A070801	Complete Loss of Power		
					CT321	Long-term surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A070801	Complete Loss of Power		
					CT384	Sensors and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070801	Complete Loss of Power		
					CT449	Generators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070801	Complete Loss of Power		
					CT981	Single use	E2403	No Clinical Signs, Symptoms or Conditions	A070801	Complete Loss of Power		
					CT983	Surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A070801	Complete Loss of Power		
					CT994	Sensors	E2403	No Clinical Signs, Symptoms or Conditions	A070801	Complete Loss of Power		
10438	2020	All other reportable incidents	47267	Single-chamber implantable pacemaker, rate-responsive	CT141	Electrodes/Leads and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	14/12/2020	27/05/2021
							F19	Surgical Intervention	A070504	Premature Discharge of Battery	14/12/2020	27/05/2021
					CT145	Cardiovascular devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	14/12/2020	27/05/2021
							F19	Surgical Intervention	A070504	Premature Discharge of Battery	14/12/2020	27/05/2021
					CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	14/12/2020	27/05/2021
							F19	Surgical Intervention	A070504	Premature Discharge of Battery	14/12/2020	27/05/2021
					CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	14/12/2020	27/05/2021
							F19	Surgical Intervention	A070504	Premature Discharge of Battery	14/12/2020	27/05/2021
					CT233	Surgical	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	14/12/2020	27/05/2021
							F19	Surgical Intervention	A070504	Premature Discharge of Battery	14/12/2020	27/05/2021
					CT2406	Implantable	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	14/12/2020	27/05/2021
							F19	Surgical Intervention	A070504	Premature Discharge of Battery	14/12/2020	27/05/2021
					CT2407	Active implantable	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	14/12/2020	27/05/2021
							F19	Surgical Intervention	A070504	Premature Discharge of Battery	14/12/2020	27/05/2021
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	14/12/2020	27/05/2021
							F19	Surgical Intervention	A070504	Premature Discharge of Battery	14/12/2020	27/05/2021
					CT2727	Active implant devices/ancillaries	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	14/12/2020	27/05/2021
							F19	Surgical Intervention	A070504	Premature Discharge of Battery	14/12/2020	27/05/2021
					CT2750	Implantable cardiac electrical stimulators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	14/12/2020	27/05/2021
							F19	Surgical Intervention	A070504	Premature Discharge of Battery	14/12/2020	27/05/2021
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	14/12/2020	27/05/2021
							F19	Surgical Intervention	A070504	Premature Discharge of Battery	14/12/2020	27/05/2021
					CT2985	Active	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	14/12/2020	27/05/2021
							F19	Surgical Intervention	A070504	Premature Discharge of Battery	14/12/2020	27/05/2021
					CT3001	Implantable metal	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	14/12/2020	27/05/2021
							F19	Surgical Intervention	A070504	Premature Discharge of Battery	14/12/2020	27/05/2021
					CT3002	Implantable synthetic polymer	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	14/12/2020	27/05/2021
							F19	Surgical Intervention	A070504	Premature Discharge of Battery	14/12/2020	27/05/2021
					CT3006	Implantable inorganic material	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	14/12/2020	27/05/2021
							F19	Surgical Intervention	A070504	Premature Discharge of Battery	14/12/2020	27/05/2021
					CT314	Electrically powered	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	14/12/2020	27/05/2021
							F19	Surgical Intervention	A070504	Premature Discharge of Battery	14/12/2020	27/05/2021
					CT321	Long-term surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	14/12/2020	27/05/2021
							F19	Surgical Intervention	A070504	Premature Discharge of Battery	14/12/2020	27/05/2021
					CT449	Generators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	14/12/2020	27/05/2021
							F19	Surgical Intervention	A070504	Premature Discharge of Battery	14/12/2020	27/05/2021

					CT981	Single use	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	14/12/2020	27/05/2021
							F19	Surgical Intervention	A070504	Premature Discharge of Battery	14/12/2020	27/05/2021
					CT983	Surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	14/12/2020	27/05/2021
							F19	Surgical Intervention	A070504	Premature Discharge of Battery	14/12/2020	27/05/2021
10439	2020	All other reportable incidents	36241	Pacemaker/defibrillator lead	CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A040101	Fracture		
									A072201	High Impedance		
							F26	No Health Consequences or Impact	A040101	Fracture		
									A072201	High Impedance		
10440	2020	All other reportable incidents	35852	Single-chamber implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A070909	Oversensing		
							F23	Unexpected Medical Intervention	A070909	Oversensing		
10441	2020	All other reportable incidents	47263	Cardiac resynchronization therapy implantable pacemaker	CT1684	Cardiac pacemakers and associated devices	E011903	Syncope/Fainting	A0712	Pacing Problem	04/12/2020	04/12/2020
									A0908	Incorrect, Inadequate or Imprecise Result or Readings	04/12/2020	04/12/2020
							E2403	No Clinical Signs, Symptoms or Conditions	A0712	Pacing Problem	04/12/2020	04/12/2020
									A0908	Incorrect, Inadequate or Imprecise Result or Readings	04/12/2020	04/12/2020
							F08	Hospitalization or Prolonged Hospitalization	A0712	Pacing Problem	04/12/2020	04/12/2020
									A0908	Incorrect, Inadequate or Imprecise Result or Readings	04/12/2020	04/12/2020
							F19	Surgical Intervention	A0712	Pacing Problem	04/12/2020	04/12/2020
									A0908	Incorrect, Inadequate or Imprecise Result or Readings	04/12/2020	04/12/2020
							F1903	Device Explantation	A0712	Pacing Problem	04/12/2020	04/12/2020
									A0908	Incorrect, Inadequate or Imprecise Result or Readings	04/12/2020	04/12/2020
							F1905	Device Revision or Replacement	A0712	Pacing Problem	04/12/2020	04/12/2020
									A0908	Incorrect, Inadequate or Imprecise Result or Readings	04/12/2020	04/12/2020
					CT1685	Cardiac pacemakers	E011903	Syncope/Fainting	A0712	Pacing Problem	04/12/2020	04/12/2020
									A0908	Incorrect, Inadequate or Imprecise Result or Readings	04/12/2020	04/12/2020
							E2403	No Clinical Signs, Symptoms or Conditions	A0712	Pacing Problem	04/12/2020	04/12/2020
									A0908	Incorrect, Inadequate or Imprecise Result or Readings	04/12/2020	04/12/2020
							F08	Hospitalization or Prolonged Hospitalization	A0712	Pacing Problem	04/12/2020	04/12/2020
									A0908	Incorrect, Inadequate or Imprecise Result or Readings	04/12/2020	04/12/2020
							F19	Surgical Intervention	A0712	Pacing Problem	04/12/2020	04/12/2020
									A0908	Incorrect, Inadequate or Imprecise Result or Readings	04/12/2020	04/12/2020
							F1903	Device Explantation	A0712	Pacing Problem	04/12/2020	04/12/2020
									A0908	Incorrect, Inadequate or Imprecise Result or Readings	04/12/2020	04/12/2020
							F1905	Device Revision or Replacement	A0712	Pacing Problem	04/12/2020	04/12/2020
									A0908	Incorrect, Inadequate or Imprecise Result or Readings	04/12/2020	04/12/2020
					CT257	Implantable cardiac pacemakers	E011903	Syncope/Fainting	A0712	Pacing Problem	04/12/2020	04/12/2020
									A0908	Incorrect, Inadequate or Imprecise Result or Readings	04/12/2020	04/12/2020
							E2403	No Clinical Signs, Symptoms or Conditions	A0712	Pacing Problem	04/12/2020	04/12/2020
									A0908	Incorrect, Inadequate or Imprecise Result or Readings	04/12/2020	04/12/2020
							F08	Hospitalization or Prolonged Hospitalization	A0712	Pacing Problem	04/12/2020	04/12/2020
									A0908	Incorrect, Inadequate or Imprecise Result or Readings	04/12/2020	04/12/2020
							F19	Surgical Intervention	A0712	Pacing Problem	04/12/2020	04/12/2020
									A0908	Incorrect, Inadequate or Imprecise Result or Readings	04/12/2020	04/12/2020
							F1903	Device Explantation	A0712	Pacing Problem	04/12/2020	04/12/2020
									A0908	Incorrect, Inadequate or Imprecise Result or Readings	04/12/2020	04/12/2020
							F1905	Device Revision or Replacement	A0712	Pacing Problem	04/12/2020	04/12/2020
									A0908	Incorrect, Inadequate or Imprecise Result or Readings	04/12/2020	04/12/2020
					CT2866	Implantable cardiac electrical stimulators	E011903	Syncope/Fainting	A0712	Pacing Problem	04/12/2020	04/12/2020
									A0908	Incorrect, Inadequate or Imprecise Result or Readings	04/12/2020	04/12/2020
							E2403	No Clinical Signs, Symptoms or Conditions	A0712	Pacing Problem	04/12/2020	04/12/2020
									A0908	Incorrect, Inadequate or Imprecise Result or Readings	04/12/2020	04/12/2020
							F08	Hospitalization or Prolonged Hospitalization	A0712	Pacing Problem	04/12/2020	04/12/2020
									A0908	Incorrect, Inadequate or Imprecise Result or Readings	04/12/2020	04/12/2020
							F19	Surgical Intervention	A0712	Pacing Problem	04/12/2020	04/12/2020
									A0908	Incorrect, Inadequate or Imprecise Result or Readings	04/12/2020	04/12/2020
							F1903	Device Explantation	A0712	Pacing Problem	04/12/2020	04/12/2020
									A0908	Incorrect, Inadequate or Imprecise Result or Readings	04/12/2020	04/12/2020
							F1905	Device Revision or Replacement	A0712	Pacing Problem	04/12/2020	04/12/2020
									A0908	Incorrect, Inadequate or Imprecise Result or Readings	04/12/2020	04/12/2020
10442	2020	All other reportable incidents	47265	Dual-chamber implantable pacemaker, rate-responsive	CT141	Electrodes/Leads and associated devices	E1020	Nausea	A070101	Failure to Capture		
									A070909	Oversensing		
							E2403	No Clinical Signs, Symptoms or Conditions	A070101	Failure to Capture		
									A070909	Oversensing		
							F26	No Health Consequences or Impact	A070101	Failure to Capture		
									A070909	Oversensing		
					CT145	Cardiovascular devices	E1020	Nausea	A070101	Failure to Capture		
									A070909	Oversensing		
							E2403	No Clinical Signs, Symptoms or Conditions	A070101	Failure to Capture		
									A070909	Oversensing		
							F26	No Health Consequences or Impact	A070101	Failure to Capture		

									A070909	Oversensing		
					CT1684	Cardiac pacemakers and associated devices	E1020	Nausea	A070101	Failure to Capture		
									A070909	Oversensing		
							E2403	No Clinical Signs, Symptoms or Conditions	A070101	Failure to Capture		
									A070909	Oversensing		
							F26	No Health Consequences or Impact	A070101	Failure to Capture		
									A070909	Oversensing		
					CT1685	Cardiac pacemakers	E1020	Nausea	A070101	Failure to Capture		
									A070909	Oversensing		
							E2403	No Clinical Signs, Symptoms or Conditions	A070101	Failure to Capture		
									A070909	Oversensing		
							F26	No Health Consequences or Impact	A070101	Failure to Capture		
									A070909	Oversensing		
					CT233	Surgical	E1020	Nausea	A070101	Failure to Capture		
									A070909	Oversensing		
							E2403	No Clinical Signs, Symptoms or Conditions	A070101	Failure to Capture		
									A070909	Oversensing		
							F26	No Health Consequences or Impact	A070101	Failure to Capture		
									A070909	Oversensing		
					CT2406	Implantable	E1020	Nausea	A070101	Failure to Capture		
									A070909	Oversensing		
							E2403	No Clinical Signs, Symptoms or Conditions	A070101	Failure to Capture		
									A070909	Oversensing		
							F26	No Health Consequences or Impact	A070101	Failure to Capture		
									A070909	Oversensing		
					CT2407	Active implantable	E1020	Nausea	A070101	Failure to Capture		
									A070909	Oversensing		
							E2403	No Clinical Signs, Symptoms or Conditions	A070101	Failure to Capture		
									A070909	Oversensing		
							F26	No Health Consequences or Impact	A070101	Failure to Capture		
									A070909	Oversensing		
					CT257	Implantable cardiac pacemakers	E1020	Nausea	A070101	Failure to Capture		
									A070909	Oversensing		
							E2403	No Clinical Signs, Symptoms or Conditions	A070101	Failure to Capture		
									A070909	Oversensing		
							F26	No Health Consequences or Impact	A070101	Failure to Capture		
									A070909	Oversensing		
					CT2727	Active implant devices/ancillaries	E1020	Nausea	A070101	Failure to Capture		
									A070909	Oversensing		
							E2403	No Clinical Signs, Symptoms or Conditions	A070101	Failure to Capture		
									A070909	Oversensing		
							F26	No Health Consequences or Impact	A070101	Failure to Capture		
									A070909	Oversensing		
					CT2750	Implantable cardiac electrical stimulators and associated devices	E1020	Nausea	A070101	Failure to Capture		
									A070909	Oversensing		
							E2403	No Clinical Signs, Symptoms or Conditions	A070101	Failure to Capture		
									A070909	Oversensing		
							F26	No Health Consequences or Impact	A070101	Failure to Capture		
									A070909	Oversensing		
					CT2866	Implantable cardiac electrical stimulators	E1020	Nausea	A070101	Failure to Capture		
									A070909	Oversensing		
							E2403	No Clinical Signs, Symptoms or Conditions	A070101	Failure to Capture		
									A070909	Oversensing		
							F26	No Health Consequences or Impact	A070101	Failure to Capture		
									A070909	Oversensing		
					CT2985	Active	E1020	Nausea	A070101	Failure to Capture		
									A070909	Oversensing		
							E2403	No Clinical Signs, Symptoms or Conditions	A070101	Failure to Capture		
									A070909	Oversensing		
							F26	No Health Consequences or Impact	A070101	Failure to Capture		
									A070909	Oversensing		
					CT3001	Implantable metal	E1020	Nausea	A070101	Failure to Capture		
									A070909	Oversensing		
							E2403	No Clinical Signs, Symptoms or Conditions	A070101	Failure to Capture		
									A070909	Oversensing		
							F26	No Health Consequences or Impact	A070101	Failure to Capture		
									A070909	Oversensing		
					CT3002	Implantable synthetic polymer	E1020	Nausea	A070101	Failure to Capture		
									A070909	Oversensing		
							E2403	No Clinical Signs, Symptoms or Conditions	A070101	Failure to Capture		
									A070909	Oversensing		
							F26	No Health Consequences or Impact	A070101	Failure to Capture		
									A070909	Oversensing		
					CT3006	Implantable inorganic material	E1020	Nausea	A070101	Failure to Capture		
									A070909	Oversensing		
							E2403	No Clinical Signs, Symptoms or Conditions	A070101	Failure to Capture		
									A070909	Oversensing		
							F26	No Health Consequences or Impact	A070101	Failure to Capture		
									A070909	Oversensing		
					CT314	Electrically powered	E1020	Nausea	A070101	Failure to Capture		
									A070909	Oversensing		
							E2403	No Clinical Signs, Symptoms or Conditions	A070101	Failure to Capture		
									A070909	Oversensing		
							F26	No Health Consequences or Impact	A070101	Failure to Capture		
									A070909	Oversensing		
					CT321	Long-term surgical invasive	E1020	Nausea	A070101	Failure to Capture		
									A070909	Oversensing		
							E2403	No Clinical Signs, Symptoms or Conditions	A070101	Failure to Capture		

									A070909	Oversensing		
							F26	No Health Consequences or Impact	A070101	Failure to Capture		
									A070909	Oversensing		
					CT449	Generators and associated devices	E1020	Nausea	A070101	Failure to Capture		
									A070909	Oversensing		
							E2403	No Clinical Signs, Symptoms or Conditions	A070101	Failure to Capture		
									A070909	Oversensing		
							F26	No Health Consequences or Impact	A070101	Failure to Capture		
									A070909	Oversensing		
					CT981	Single use	E1020	Nausea	A070101	Failure to Capture		
									A070909	Oversensing		
							E2403	No Clinical Signs, Symptoms or Conditions	A070101	Failure to Capture		
									A070909	Oversensing		
							F26	No Health Consequences or Impact	A070101	Failure to Capture		
									A070909	Oversensing		
					CT983	Surgical invasive	E1020	Nausea	A070101	Failure to Capture		
									A070909	Oversensing		
							E2403	No Clinical Signs, Symptoms or Conditions	A070101	Failure to Capture		
									A070909	Oversensing		
							F26	No Health Consequences or Impact	A070101	Failure to Capture		
									A070909	Oversensing		
10443	2020	All other reportable incidents	35852	Single-chamber implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A070909	Oversensing		
							F26	No Health Consequences or Impact	A070909	Oversensing		
10444	2020	All other reportable incidents	35852	Single-chamber implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E2104	Electric Shock	A071302	Inappropriate/Inadequate Shock/Stimulation		
							F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A071302	Inappropriate/Inadequate Shock/Stimulation		
							F26	No Health Consequences or Impact	A071302	Inappropriate/Inadequate Shock/Stimulation		
10445	2020	All other reportable incidents	36241	Pacemaker/defibrillator lead	CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A05	Mechanical Problem	07/12/2020	07/12/2020
									A051201	Device Dislodged or Dislocated	07/12/2020	07/12/2020
							F08	Hospitalization or Prolonged Hospitalization	A05	Mechanical Problem	07/12/2020	07/12/2020
									A051201	Device Dislodged or Dislocated	07/12/2020	07/12/2020
							F12	Serious Injury/ Illness/ Impairment	A05	Mechanical Problem	07/12/2020	07/12/2020
									A051201	Device Dislodged or Dislocated	07/12/2020	07/12/2020
							F1903	Device Explantation	A05	Mechanical Problem	07/12/2020	07/12/2020
									A051201	Device Dislodged or Dislocated	07/12/2020	07/12/2020
							F1905	Device Revision or Replacement	A05	Mechanical Problem	07/12/2020	07/12/2020
									A051201	Device Dislodged or Dislocated	07/12/2020	07/12/2020
10446	2020	All other reportable incidents	17882	Physiologic-monitoring defibrillation system	CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A0719	Unexpected Shutdown		
							F27	Problem Identified During Non-Clinical Procedure	A0719	Unexpected Shutdown		
10447	2020	All other reportable incidents	17882	Physiologic-monitoring defibrillation system	CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A0719	Unexpected Shutdown		
							F27	Problem Identified During Non-Clinical Procedure	A0719	Unexpected Shutdown		
10448	2020	All other reportable incidents	17882	Physiologic-monitoring defibrillation system	CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A0719	Unexpected Shutdown		
							F27	Problem Identified During Non-Clinical Procedure	A0719	Unexpected Shutdown		
10449	2020	All other reportable incidents	17882	Physiologic-monitoring defibrillation system	CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A0719	Unexpected Shutdown		
							F27	Problem Identified During Non-Clinical Procedure	A0719	Unexpected Shutdown		
10450	2020	All other reportable incidents	17882	Physiologic-monitoring defibrillation system	CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A0719	Unexpected Shutdown		
							F26	No Health Consequences or Impact	A0719	Unexpected Shutdown		
10451	2020	All other reportable incidents	17882	Physiologic-monitoring defibrillation system	CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A071303	Intermittent Shock/Stimulation		
							F27	Problem Identified During Non-Clinical Procedure	A071303	Intermittent Shock/Stimulation		
10452	2020	All other reportable incidents	47263	Cardiac resynchronization therapy implantable pacemaker	CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A072201	High Impedance		
							F26	No Health Consequences or Impact	A072201	High Impedance		
					CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A072201	High Impedance		
							F26	No Health Consequences or Impact	A072201	High Impedance		
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A072201	High Impedance		
							F26	No Health Consequences or Impact	A072201	High Impedance		
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A072201	High Impedance		
							F26	No Health Consequences or Impact	A072201	High Impedance		
10453	2020	All other reportable incidents	37265	Dual-chamber implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E2104	Electric Shock	A070910	Undersensing		
									A0712	Pacing Problem		
									A071302	Inappropriate/Inadequate Shock/Stimulation		
							F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A070910	Undersensing		
									A0712	Pacing Problem		
									A071302	Inappropriate/Inadequate Shock/Stimulation		
							F26	No Health Consequences or Impact	A070910	Undersensing		
									A0712	Pacing Problem		
									A071302	Inappropriate/Inadequate Shock/Stimulation		
10454	2020	All other reportable incidents	36241	Pacemaker/defibrillator lead	CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A040101	Fracture	16/10/2020	16/10/2020
									A072201	High Impedance	16/10/2020	16/10/2020
							F08	Hospitalization or Prolonged Hospitalization	A040101	Fracture	16/10/2020	16/10/2020
									A072201	High Impedance	16/10/2020	16/10/2020
							F12	Serious Injury/ Illness/ Impairment	A040101	Fracture	16/10/2020	16/10/2020
									A072201	High Impedance	16/10/2020	16/10/2020
							F19	Surgical Intervention	A040101	Fracture	16/10/2020	16/10/2020
									A072201	High Impedance	16/10/2020	16/10/2020
10455	2020	All other reportable incidents	47265	Dual-chamber implantable pacemaker, rate-responsive	CT141	Electrodes/Leads and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem	13/01/2021	13/01/2021
									A070504	Premature Discharge of Battery	13/01/2021	13/01/2021
							F19	Surgical Intervention	A0705	Battery Problem	13/01/2021	13/01/2021
									A070504	Premature Discharge of Battery	13/01/2021	13/01/2021
					CT145	Cardiovascular devices	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem	13/01/2021	13/01/2021
									A070504	Premature Discharge of Battery	13/01/2021	13/01/2021
							F19	Surgical Intervention	A0705	Battery Problem	13/01/2021	13/01/2021
									A070504	Premature Discharge of Battery	13/01/2021	13/01/2021
					CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem	13/01/2021	13/01/2021
									A070504	Premature Discharge of Battery	13/01/2021	13/01/2021
							F19	Surgical Intervention	A0705	Battery Problem	13/01/2021	13/01/2021
									A070504	Premature Discharge of Battery	13/01/2021	13/01/2021
					CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem	13/01/2021	13/01/2021
									A070504	Premature Discharge of Battery	13/01/2021	13/01/2021
							F19	Surgical Intervention	A0705	Battery Problem	13/01/2021	13/01/2021

									A070504	Premature Discharge of Battery	13/01/2021	13/01/2021
					CT233	Surgical	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem	13/01/2021	13/01/2021
									A070504	Premature Discharge of Battery	13/01/2021	13/01/2021
							F19	Surgical Intervention	A0705	Battery Problem	13/01/2021	13/01/2021
									A070504	Premature Discharge of Battery	13/01/2021	13/01/2021
					CT2406	Implantable	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem	13/01/2021	13/01/2021
									A070504	Premature Discharge of Battery	13/01/2021	13/01/2021
							F19	Surgical Intervention	A0705	Battery Problem	13/01/2021	13/01/2021
									A070504	Premature Discharge of Battery	13/01/2021	13/01/2021
					CT2407	Active implantable	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem	13/01/2021	13/01/2021
									A070504	Premature Discharge of Battery	13/01/2021	13/01/2021
							F19	Surgical Intervention	A0705	Battery Problem	13/01/2021	13/01/2021
									A070504	Premature Discharge of Battery	13/01/2021	13/01/2021
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem	13/01/2021	13/01/2021
									A070504	Premature Discharge of Battery	13/01/2021	13/01/2021
							F19	Surgical Intervention	A0705	Battery Problem	13/01/2021	13/01/2021
									A070504	Premature Discharge of Battery	13/01/2021	13/01/2021
					CT2727	Active implant devices/ancillaries	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem	13/01/2021	13/01/2021
									A070504	Premature Discharge of Battery	13/01/2021	13/01/2021
							F19	Surgical Intervention	A0705	Battery Problem	13/01/2021	13/01/2021
									A070504	Premature Discharge of Battery	13/01/2021	13/01/2021
					CT2750	Implantable cardiac electrical stimulators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem	13/01/2021	13/01/2021
									A070504	Premature Discharge of Battery	13/01/2021	13/01/2021
							F19	Surgical Intervention	A0705	Battery Problem	13/01/2021	13/01/2021
									A070504	Premature Discharge of Battery	13/01/2021	13/01/2021
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem	13/01/2021	13/01/2021
									A070504	Premature Discharge of Battery	13/01/2021	13/01/2021
							F19	Surgical Intervention	A0705	Battery Problem	13/01/2021	13/01/2021
									A070504	Premature Discharge of Battery	13/01/2021	13/01/2021
					CT2985	Active	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem	13/01/2021	13/01/2021
									A070504	Premature Discharge of Battery	13/01/2021	13/01/2021
							F19	Surgical Intervention	A0705	Battery Problem	13/01/2021	13/01/2021
									A070504	Premature Discharge of Battery	13/01/2021	13/01/2021
					CT3001	Implantable metal	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem	13/01/2021	13/01/2021
									A070504	Premature Discharge of Battery	13/01/2021	13/01/2021
							F19	Surgical Intervention	A0705	Battery Problem	13/01/2021	13/01/2021
									A070504	Premature Discharge of Battery	13/01/2021	13/01/2021
					CT3002	Implantable synthetic polymer	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem	13/01/2021	13/01/2021
									A070504	Premature Discharge of Battery	13/01/2021	13/01/2021
							F19	Surgical Intervention	A0705	Battery Problem	13/01/2021	13/01/2021
									A070504	Premature Discharge of Battery	13/01/2021	13/01/2021
					CT3006	Implantable inorganic material	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem	13/01/2021	13/01/2021
									A070504	Premature Discharge of Battery	13/01/2021	13/01/2021
							F19	Surgical Intervention	A0705	Battery Problem	13/01/2021	13/01/2021
									A070504	Premature Discharge of Battery	13/01/2021	13/01/2021
					CT314	Electrically powered	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem	13/01/2021	13/01/2021
									A070504	Premature Discharge of Battery	13/01/2021	13/01/2021
							F19	Surgical Intervention	A0705	Battery Problem	13/01/2021	13/01/2021
									A070504	Premature Discharge of Battery	13/01/2021	13/01/2021
					CT321	Long-term surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem	13/01/2021	13/01/2021
									A070504	Premature Discharge of Battery	13/01/2021	13/01/2021
							F19	Surgical Intervention	A0705	Battery Problem	13/01/2021	13/01/2021
									A070504	Premature Discharge of Battery	13/01/2021	13/01/2021
					CT449	Generators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem	13/01/2021	13/01/2021
									A070504	Premature Discharge of Battery	13/01/2021	13/01/2021
							F19	Surgical Intervention	A0705	Battery Problem	13/01/2021	13/01/2021
									A070504	Premature Discharge of Battery	13/01/2021	13/01/2021
					CT981	Single use	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem	13/01/2021	13/01/2021
									A070504	Premature Discharge of Battery	13/01/2021	13/01/2021
							F19	Surgical Intervention	A0705	Battery Problem	13/01/2021	13/01/2021
									A070504	Premature Discharge of Battery	13/01/2021	13/01/2021
					CT983	Surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem	13/01/2021	13/01/2021
									A070504	Premature Discharge of Battery	13/01/2021	13/01/2021
							F19	Surgical Intervention	A0705	Battery Problem	13/01/2021	13/01/2021
									A070504	Premature Discharge of Battery	13/01/2021	13/01/2021
10456	2020	All other reportable incidents	36241	Pacemaker/defibrillator lead	CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A0203	Defective Device	16/10/2020	16/10/2020
							F08	Hospitalization or Prolonged Hospitalization	A0203	Defective Device	16/10/2020	16/10/2020
							F12	Serious Injury/ Illness/ Impairment	A0203	Defective Device	16/10/2020	16/10/2020
							F19	Surgical Intervention	A0203	Defective Device	16/10/2020	16/10/2020
							F1905	Device Revision or Replacement	A0203	Defective Device	16/10/2020	16/10/2020
10457	2020	All other reportable incidents	36241	Pacemaker/defibrillator lead	CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A05	Mechanical Problem	16/10/2020	16/10/2020
									A072201	High Impedance	16/10/2020	16/10/2020
							F1908	Prolonged Surgery	A05	Mechanical Problem	16/10/2020	16/10/2020
									A072201	High Impedance	16/10/2020	16/10/2020
10458	2020	All other reportable incidents	35852	Single-chamber implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A070909	Oversensing		
							F26	No Health Consequences or Impact	A070909	Oversensing		
10459	2020	All other reportable incidents	35852	Single-chamber implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	25/01/2021	25/01/2021
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	25/01/2021	25/01/2021
10460	2020	All other reportable incidents	35223	Endocardial/interventricular septal pacing lead	CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070901	Decreased Sensitivity		
									A0722	Impedance Problem		
							F26	No Health Consequences or Impact	A070901	Decreased Sensitivity		
									A0722	Impedance Problem		
10461	2020	All other reportable incidents	35852	Single-chamber implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	20/02/2021	20/02/2021
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	20/02/2021	20/02/2021
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	20/02/2021	20/02/2021
10462	2020	All other reportable incidents	17882	Physiologic-monitoring defibrillation system	CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A090811	Unable to Obtain Readings		
							F26	No Health Consequences or Impact	A090811	Unable to Obtain Readings		
10463	2021	All other reportable incidents	35852	Single-chamber implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	13/01/2021	13/01/2021

							F19	Surgical Intervention	A070504	Premature Discharge of Battery	13/01/2021	13/01/2021
10464	2021	All other reportable incidents	35223	Endocardial/interventricular septal pacing lead	CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A0401	Break		
									A070101	Failure to Capture		
									A070103	Intermittent Capture		
									A090801	Signal Artifact/Noise		
							F26	No Health Consequences or Impact	A0401	Break		
									A070101	Failure to Capture		
									A070103	Intermittent Capture		
									A090801	Signal Artifact/Noise		
10465	2021	All other reportable incidents	35223	Endocardial/interventricular septal pacing lead	CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A0401	Break		
									A070909	Oversensing		
							F26	No Health Consequences or Impact	A0401	Break		
									A070909	Oversensing		
10466	2021	All other reportable incidents	35223	Endocardial/interventricular septal pacing lead	CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A0401	Break		
									A070909	Oversensing		
							F26	No Health Consequences or Impact	A0401	Break		
									A070909	Oversensing		
10467	2021	All other reportable incidents	35223	Endocardial/interventricular septal pacing lead	CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A07	Electrical /Electronic Property Problem		
									A072201	High Impedance		
							F26	No Health Consequences or Impact	A07	Electrical /Electronic Property Problem		
									A072201	High Impedance		
10468	2021	All other reportable incidents	47270	Cardiac resynchronization therapy implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A070912	Incorrect Interpretation of Signal		
									A071302	Inappropriate/Inadequate Shock/Stimulation		
							F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A070912	Incorrect Interpretation of Signal		
									A071302	Inappropriate/Inadequate Shock/Stimulation		
							F23	Unexpected Medical Intervention	A070912	Incorrect Interpretation of Signal		
									A071302	Inappropriate/Inadequate Shock/Stimulation		
10469	2021	All other reportable incidents	47270	Cardiac resynchronization therapy implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
10470	2021	All other reportable incidents	36241	Pacemaker/defibrillator lead	CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A072201	High Impedance		
							F26	No Health Consequences or Impact	A072201	High Impedance		
10471	2021	All other reportable incidents	47270	Cardiac resynchronization therapy implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A072201	High Impedance		
							F26	No Health Consequences or Impact	A072201	High Impedance		
10472	2021	All other reportable incidents	47265	Dual-chamber implantable pacemaker, rate-responsive	CT141	Electrodes/Leads and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
									A090804	False Positive Result		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
									A090804	False Positive Result		
					CT145	Cardiovascular devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
									A090804	False Positive Result		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
									A090804	False Positive Result		
					CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
									A090804	False Positive Result		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
									A090804	False Positive Result		
					CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
									A090804	False Positive Result		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
									A090804	False Positive Result		
					CT233	Surgical	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
									A090804	False Positive Result		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
									A090804	False Positive Result		
					CT2406	Implantable	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
									A090804	False Positive Result		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
									A090804	False Positive Result		
					CT2407	Active implantable	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
									A090804	False Positive Result		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
									A090804	False Positive Result		
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
									A090804	False Positive Result		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
									A090804	False Positive Result		
					CT2727	Active implant devices/ancillaries	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
									A090804	False Positive Result		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
									A090804	False Positive Result		
					CT2750	Implantable cardiac electrical stimulators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
									A090804	False Positive Result		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
									A090804	False Positive Result		
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
									A090804	False Positive Result		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
									A090804	False Positive Result		
					CT2985	Active	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
									A090804	False Positive Result		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
									A090804	False Positive Result		
					CT3001	Implantable metal	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
									A090804	False Positive Result		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
									A090804	False Positive Result		
					CT3002	Implantable synthetic polymer	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
									A090804	False Positive Result		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		

								A090804	False Positive Result			
				CT3006	Implantable inorganic material	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery			
								A090804	False Positive Result			
						F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery			
								A090804	False Positive Result			
				CT314	Electrically powered	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery			
								A090804	False Positive Result			
						F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery			
								A090804	False Positive Result			
				CT321	Long-term surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery			
								A090804	False Positive Result			
						F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery			
								A090804	False Positive Result			
				CT449	Generators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery			
								A090804	False Positive Result			
						F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery			
								A090804	False Positive Result			
				CT981	Single use	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery			
								A090804	False Positive Result			
						F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery			
								A090804	False Positive Result			
				CT983	Surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery			
								A090804	False Positive Result			
						F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery			
								A090804	False Positive Result			
10473	2021	All other reportable incidents	36241	Pacemaker/defibrillator lead	CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A072201	High Impedance		
							F15	Recognized Device or Procedural Complication	A072201	High Impedance		
10474	2021	All other reportable incidents	47263	Cardiac resynchronization therapy implantable pacemaker	CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	08/01/2021	08/01/2021
							F1901	Additional Surgery	A070504	Premature Discharge of Battery	08/01/2021	08/01/2021
					CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	08/01/2021	08/01/2021
							F1901	Additional Surgery	A070504	Premature Discharge of Battery	08/01/2021	08/01/2021
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	08/01/2021	08/01/2021
							F1901	Additional Surgery	A070504	Premature Discharge of Battery	08/01/2021	08/01/2021
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	08/01/2021	08/01/2021
							F1901	Additional Surgery	A070504	Premature Discharge of Battery	08/01/2021	08/01/2021
10475	2021	Death	35852	Single-chamber implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E0602	Cardiac Arrest	A070910	Undersensing		
							F02	Death	A070910	Undersensing		
10476	2021	All other reportable incidents	36241	Pacemaker/defibrillator lead	CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070909	Oversensing	11/12/2020	11/12/2020
									A071302	Inappropriate/Inadequate Shock/Stimulation	11/12/2020	11/12/2020
									A090801	Signal Artifact/Noise	11/12/2020	11/12/2020
									A1301	Failure to Read Input Signal	11/12/2020	11/12/2020
							F08	Hospitalization or Prolonged Hospitalization	A070909	Oversensing	11/12/2020	11/12/2020
									A071302	Inappropriate/Inadequate Shock/Stimulation	11/12/2020	11/12/2020
									A090801	Signal Artifact/Noise	11/12/2020	11/12/2020
									A1301	Failure to Read Input Signal	11/12/2020	11/12/2020
							F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A070909	Oversensing	11/12/2020	11/12/2020
									A071302	Inappropriate/Inadequate Shock/Stimulation	11/12/2020	11/12/2020
									A090801	Signal Artifact/Noise	11/12/2020	11/12/2020
									A1301	Failure to Read Input Signal	11/12/2020	11/12/2020
							F12	Serious Injury/ Illness/ Impairment	A070909	Oversensing	11/12/2020	11/12/2020
									A071302	Inappropriate/Inadequate Shock/Stimulation	11/12/2020	11/12/2020
									A090801	Signal Artifact/Noise	11/12/2020	11/12/2020
									A1301	Failure to Read Input Signal	11/12/2020	11/12/2020
							F19	Surgical Intervention	A070909	Oversensing	11/12/2020	11/12/2020
									A071302	Inappropriate/Inadequate Shock/Stimulation	11/12/2020	11/12/2020
									A090801	Signal Artifact/Noise	11/12/2020	11/12/2020
									A1301	Failure to Read Input Signal	11/12/2020	11/12/2020
							F1905	Device Revision or Replacement	A070909	Oversensing	11/12/2020	11/12/2020
									A071302	Inappropriate/Inadequate Shock/Stimulation	11/12/2020	11/12/2020
									A090801	Signal Artifact/Noise	11/12/2020	11/12/2020
									A1301	Failure to Read Input Signal	11/12/2020	11/12/2020
10477	2021	All other reportable incidents	36241	Pacemaker/defibrillator lead	CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070909	Oversensing		
									A072201	High Impedance		
							F26	No Health Consequences or Impact	A070909	Oversensing		
									A072201	High Impedance		
10478	2021	All other reportable incidents	35223	Endocardial/interventricular septal pacing lead	CT1684	Cardiac pacemakers and associated devices	E0112	Dizziness	A05	Mechanical Problem		
									A070909	Oversensing		
									A0712	Pacing Problem		
							F23	Unexpected Medical Intervention	A05	Mechanical Problem		
									A070909	Oversensing		
									A0712	Pacing Problem		
10479	2021	All other reportable incidents	47267	Single-chamber implantable pacemaker, rate-responsive	CT141	Electrodes/Leads and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem		
							F27	Problem Identified During Non-Clinical Procedure	A0705	Battery Problem		
					CT145	Cardiovascular devices	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem		
							F27	Problem Identified During Non-Clinical Procedure	A0705	Battery Problem		
					CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem		
							F27	Problem Identified During Non-Clinical Procedure	A0705	Battery Problem		
					CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem		
							F27	Problem Identified During Non-Clinical Procedure	A0705	Battery Problem		
					CT233	Surgical	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem		
							F27	Problem Identified During Non-Clinical Procedure	A0705	Battery Problem		
					CT2406	Implantable	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem		
							F27	Problem Identified During Non-Clinical Procedure	A0705	Battery Problem		
					CT2407	Active implantable	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem		
							F27	Problem Identified During Non-Clinical Procedure	A0705	Battery Problem		
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem		
							F27	Problem Identified During Non-Clinical Procedure	A0705	Battery Problem		
					CT2727	Active implant devices/ancillaries	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem		

						F27	Problem Identified During Non-Clinical Procedure	A0705	Battery Problem			
				CT2750	Implantable cardiac electrical stimulators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem			
						F27	Problem Identified During Non-Clinical Procedure	A0705	Battery Problem			
				CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem			
						F27	Problem Identified During Non-Clinical Procedure	A0705	Battery Problem			
				CT2985	Active	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem			
						F27	Problem Identified During Non-Clinical Procedure	A0705	Battery Problem			
				CT3001	Implantable metal	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem			
						F27	Problem Identified During Non-Clinical Procedure	A0705	Battery Problem			
				CT3002	Implantable synthetic polymer	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem			
						F27	Problem Identified During Non-Clinical Procedure	A0705	Battery Problem			
				CT3006	Implantable inorganic material	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem			
						F27	Problem Identified During Non-Clinical Procedure	A0705	Battery Problem			
				CT314	Electrically powered	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem			
						F27	Problem Identified During Non-Clinical Procedure	A0705	Battery Problem			
				CT321	Long-term surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem			
						F27	Problem Identified During Non-Clinical Procedure	A0705	Battery Problem			
				CT449	Generators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem			
						F27	Problem Identified During Non-Clinical Procedure	A0705	Battery Problem			
				CT981	Single use	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem			
						F27	Problem Identified During Non-Clinical Procedure	A0705	Battery Problem			
				CT983	Surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem			
						F27	Problem Identified During Non-Clinical Procedure	A0705	Battery Problem			
10480	2021	All other reportable incidents	47265	Dual-chamber implantable pacemaker, rate-responsive	CT141	Electrodes/Leads and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	31/01/2021	31/01/2021
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	31/01/2021	31/01/2021
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	31/01/2021	31/01/2021
					CT145	Cardiovascular devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	31/01/2021	31/01/2021
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	31/01/2021	31/01/2021
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	31/01/2021	31/01/2021
					CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	31/01/2021	31/01/2021
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	31/01/2021	31/01/2021
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	31/01/2021	31/01/2021
					CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	31/01/2021	31/01/2021
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	31/01/2021	31/01/2021
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	31/01/2021	31/01/2021
					CT233	Surgical	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	31/01/2021	31/01/2021
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	31/01/2021	31/01/2021
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	31/01/2021	31/01/2021
					CT2406	Implantable	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	31/01/2021	31/01/2021
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	31/01/2021	31/01/2021
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	31/01/2021	31/01/2021
					CT2407	Active implantable	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	31/01/2021	31/01/2021
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	31/01/2021	31/01/2021
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	31/01/2021	31/01/2021
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	31/01/2021	31/01/2021
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	31/01/2021	31/01/2021
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	31/01/2021	31/01/2021
					CT2727	Active implant devices/ancillaries	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	31/01/2021	31/01/2021
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	31/01/2021	31/01/2021
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	31/01/2021	31/01/2021
					CT2750	Implantable cardiac electrical stimulators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	31/01/2021	31/01/2021
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	31/01/2021	31/01/2021
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	31/01/2021	31/01/2021
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	31/01/2021	31/01/2021
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	31/01/2021	31/01/2021
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	31/01/2021	31/01/2021
					CT2985	Active	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	31/01/2021	31/01/2021
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	31/01/2021	31/01/2021
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	31/01/2021	31/01/2021
					CT3001	Implantable metal	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	31/01/2021	31/01/2021
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	31/01/2021	31/01/2021
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	31/01/2021	31/01/2021
					CT3002	Implantable synthetic polymer	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	31/01/2021	31/01/2021
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	31/01/2021	31/01/2021
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	31/01/2021	31/01/2021
					CT3006	Implantable inorganic material	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	31/01/2021	31/01/2021
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	31/01/2021	31/01/2021
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	31/01/2021	31/01/2021
					CT314	Electrically powered	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	31/01/2021	31/01/2021
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	31/01/2021	31/01/2021
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	31/01/2021	31/01/2021
					CT321	Long-term surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	31/01/2021	31/01/2021
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	31/01/2021	31/01/2021
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	31/01/2021	31/01/2021
					CT449	Generators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	31/01/2021	31/01/2021
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	31/01/2021	31/01/2021
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	31/01/2021	31/01/2021
					CT981	Single use	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	31/01/2021	31/01/2021
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	31/01/2021	31/01/2021
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	31/01/2021	31/01/2021
					CT983	Surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	31/01/2021	31/01/2021
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	31/01/2021	31/01/2021
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	31/01/2021	31/01/2021
10481	2021	All other reportable incidents	35852	Single-chamber implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A070909	Oversensing		
							F26	No Health Consequences or Impact	A070909	Oversensing		
10482	2021	All other reportable incidents	47267	Single-chamber implantable pacemaker, rate-responsive	CT141	Electrodes/Leads and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A0908	Incorrect, Inadequate or Imprecise Result or Readings	09/01/2021	09/01/2021

							F26	No Health Consequences or Impact	A0908	Incorrect, Inadequate or Imprecise Result or Readings	09/01/2021	09/01/2021
					CT145	Cardiovascular devices	E2403	No Clinical Signs, Symptoms or Conditions	A0908	Incorrect, Inadequate or Imprecise Result or Readings	09/01/2021	09/01/2021
							F26	No Health Consequences or Impact	A0908	Incorrect, Inadequate or Imprecise Result or Readings	09/01/2021	09/01/2021
					CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A0908	Incorrect, Inadequate or Imprecise Result or Readings	09/01/2021	09/01/2021
							F26	No Health Consequences or Impact	A0908	Incorrect, Inadequate or Imprecise Result or Readings	09/01/2021	09/01/2021
					CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A0908	Incorrect, Inadequate or Imprecise Result or Readings	09/01/2021	09/01/2021
							F26	No Health Consequences or Impact	A0908	Incorrect, Inadequate or Imprecise Result or Readings	09/01/2021	09/01/2021
					CT233	Surgical	E2403	No Clinical Signs, Symptoms or Conditions	A0908	Incorrect, Inadequate or Imprecise Result or Readings	09/01/2021	09/01/2021
							F26	No Health Consequences or Impact	A0908	Incorrect, Inadequate or Imprecise Result or Readings	09/01/2021	09/01/2021
					CT2406	Implantable	E2403	No Clinical Signs, Symptoms or Conditions	A0908	Incorrect, Inadequate or Imprecise Result or Readings	09/01/2021	09/01/2021
							F26	No Health Consequences or Impact	A0908	Incorrect, Inadequate or Imprecise Result or Readings	09/01/2021	09/01/2021
					CT2407	Active implantable	E2403	No Clinical Signs, Symptoms or Conditions	A0908	Incorrect, Inadequate or Imprecise Result or Readings	09/01/2021	09/01/2021
							F26	No Health Consequences or Impact	A0908	Incorrect, Inadequate or Imprecise Result or Readings	09/01/2021	09/01/2021
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A0908	Incorrect, Inadequate or Imprecise Result or Readings	09/01/2021	09/01/2021
							F26	No Health Consequences or Impact	A0908	Incorrect, Inadequate or Imprecise Result or Readings	09/01/2021	09/01/2021
					CT2727	Active implant devices/ancillaries	E2403	No Clinical Signs, Symptoms or Conditions	A0908	Incorrect, Inadequate or Imprecise Result or Readings	09/01/2021	09/01/2021
							F26	No Health Consequences or Impact	A0908	Incorrect, Inadequate or Imprecise Result or Readings	09/01/2021	09/01/2021
					CT2750	Implantable cardiac electrical stimulators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A0908	Incorrect, Inadequate or Imprecise Result or Readings	09/01/2021	09/01/2021
							F26	No Health Consequences or Impact	A0908	Incorrect, Inadequate or Imprecise Result or Readings	09/01/2021	09/01/2021
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A0908	Incorrect, Inadequate or Imprecise Result or Readings	09/01/2021	09/01/2021
							F26	No Health Consequences or Impact	A0908	Incorrect, Inadequate or Imprecise Result or Readings	09/01/2021	09/01/2021
					CT2985	Active	E2403	No Clinical Signs, Symptoms or Conditions	A0908	Incorrect, Inadequate or Imprecise Result or Readings	09/01/2021	09/01/2021
							F26	No Health Consequences or Impact	A0908	Incorrect, Inadequate or Imprecise Result or Readings	09/01/2021	09/01/2021
					CT3001	Implantable metal	E2403	No Clinical Signs, Symptoms or Conditions	A0908	Incorrect, Inadequate or Imprecise Result or Readings	09/01/2021	09/01/2021
							F26	No Health Consequences or Impact	A0908	Incorrect, Inadequate or Imprecise Result or Readings	09/01/2021	09/01/2021
					CT3002	Implantable synthetic polymer	E2403	No Clinical Signs, Symptoms or Conditions	A0908	Incorrect, Inadequate or Imprecise Result or Readings	09/01/2021	09/01/2021
							F26	No Health Consequences or Impact	A0908	Incorrect, Inadequate or Imprecise Result or Readings	09/01/2021	09/01/2021
					CT3006	Implantable inorganic material	E2403	No Clinical Signs, Symptoms or Conditions	A0908	Incorrect, Inadequate or Imprecise Result or Readings	09/01/2021	09/01/2021
							F26	No Health Consequences or Impact	A0908	Incorrect, Inadequate or Imprecise Result or Readings	09/01/2021	09/01/2021
					CT314	Electrically powered	E2403	No Clinical Signs, Symptoms or Conditions	A0908	Incorrect, Inadequate or Imprecise Result or Readings	09/01/2021	09/01/2021
							F26	No Health Consequences or Impact	A0908	Incorrect, Inadequate or Imprecise Result or Readings	09/01/2021	09/01/2021
					CT321	Long-term surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A0908	Incorrect, Inadequate or Imprecise Result or Readings	09/01/2021	09/01/2021
							F26	No Health Consequences or Impact	A0908	Incorrect, Inadequate or Imprecise Result or Readings	09/01/2021	09/01/2021
					CT449	Generators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A0908	Incorrect, Inadequate or Imprecise Result or Readings	09/01/2021	09/01/2021
							F26	No Health Consequences or Impact	A0908	Incorrect, Inadequate or Imprecise Result or Readings	09/01/2021	09/01/2021
					CT981	Single use	E2403	No Clinical Signs, Symptoms or Conditions	A0908	Incorrect, Inadequate or Imprecise Result or Readings	09/01/2021	09/01/2021
							F26	No Health Consequences or Impact	A0908	Incorrect, Inadequate or Imprecise Result or Readings	09/01/2021	09/01/2021
					CT983	Surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A0908	Incorrect, Inadequate or Imprecise Result or Readings	09/01/2021	09/01/2021
							F26	No Health Consequences or Impact	A0908	Incorrect, Inadequate or Imprecise Result or Readings	09/01/2021	09/01/2021
10483	2021	All other reportable incidents	47267	Single-chamber implantable pacemaker, rate-responsive	CT141	Electrodes/Leads and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem	03/03/2021	03/03/2021
									A070504	Premature Discharge of Battery	03/03/2021	03/03/2021
							F08	Hospitalization or Prolonged Hospitalization	A0705	Battery Problem	03/03/2021	03/03/2021
									A070504	Premature Discharge of Battery	03/03/2021	03/03/2021
							F19	Surgical Intervention	A0705	Battery Problem	03/03/2021	03/03/2021
									A070504	Premature Discharge of Battery	03/03/2021	03/03/2021
							F1905	Device Revision or Replacement	A0705	Battery Problem	03/03/2021	03/03/2021
									A070504	Premature Discharge of Battery	03/03/2021	03/03/2021
					CT145	Cardiovascular devices	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem	03/03/2021	03/03/2021
									A070504	Premature Discharge of Battery	03/03/2021	03/03/2021
							F08	Hospitalization or Prolonged Hospitalization	A0705	Battery Problem	03/03/2021	03/03/2021
									A070504	Premature Discharge of Battery	03/03/2021	03/03/2021
							F19	Surgical Intervention	A0705	Battery Problem	03/03/2021	03/03/2021
									A070504	Premature Discharge of Battery	03/03/2021	03/03/2021
							F1905	Device Revision or Replacement	A0705	Battery Problem	03/03/2021	03/03/2021
									A070504	Premature Discharge of Battery	03/03/2021	03/03/2021
					CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem	03/03/2021	03/03/2021
									A070504	Premature Discharge of Battery	03/03/2021	03/03/2021
							F08	Hospitalization or Prolonged Hospitalization	A0705	Battery Problem	03/03/2021	03/03/2021

									A070504	Premature Discharge of Battery	03/03/2021	03/03/2021
							F19	Surgical Intervention	A0705	Battery Problem	03/03/2021	03/03/2021
									A070504	Premature Discharge of Battery	03/03/2021	03/03/2021
							F1905	Device Revision or Replacement	A0705	Battery Problem	03/03/2021	03/03/2021
									A070504	Premature Discharge of Battery	03/03/2021	03/03/2021
					CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem	03/03/2021	03/03/2021
									A070504	Premature Discharge of Battery	03/03/2021	03/03/2021
							F08	Hospitalization or Prolonged Hospitalization	A0705	Battery Problem	03/03/2021	03/03/2021
									A070504	Premature Discharge of Battery	03/03/2021	03/03/2021
							F19	Surgical Intervention	A0705	Battery Problem	03/03/2021	03/03/2021
									A070504	Premature Discharge of Battery	03/03/2021	03/03/2021
							F1905	Device Revision or Replacement	A0705	Battery Problem	03/03/2021	03/03/2021
									A070504	Premature Discharge of Battery	03/03/2021	03/03/2021
					CT233	Surgical	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem	03/03/2021	03/03/2021
									A070504	Premature Discharge of Battery	03/03/2021	03/03/2021
							F08	Hospitalization or Prolonged Hospitalization	A0705	Battery Problem	03/03/2021	03/03/2021
									A070504	Premature Discharge of Battery	03/03/2021	03/03/2021
							F19	Surgical Intervention	A0705	Battery Problem	03/03/2021	03/03/2021
									A070504	Premature Discharge of Battery	03/03/2021	03/03/2021
							F1905	Device Revision or Replacement	A0705	Battery Problem	03/03/2021	03/03/2021
									A070504	Premature Discharge of Battery	03/03/2021	03/03/2021
					CT2406	Implantable	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem	03/03/2021	03/03/2021
									A070504	Premature Discharge of Battery	03/03/2021	03/03/2021
							F08	Hospitalization or Prolonged Hospitalization	A0705	Battery Problem	03/03/2021	03/03/2021
									A070504	Premature Discharge of Battery	03/03/2021	03/03/2021
							F19	Surgical Intervention	A0705	Battery Problem	03/03/2021	03/03/2021
									A070504	Premature Discharge of Battery	03/03/2021	03/03/2021
							F1905	Device Revision or Replacement	A0705	Battery Problem	03/03/2021	03/03/2021
									A070504	Premature Discharge of Battery	03/03/2021	03/03/2021
					CT2407	Active implantable	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem	03/03/2021	03/03/2021
									A070504	Premature Discharge of Battery	03/03/2021	03/03/2021
							F08	Hospitalization or Prolonged Hospitalization	A0705	Battery Problem	03/03/2021	03/03/2021
									A070504	Premature Discharge of Battery	03/03/2021	03/03/2021
							F19	Surgical Intervention	A0705	Battery Problem	03/03/2021	03/03/2021
									A070504	Premature Discharge of Battery	03/03/2021	03/03/2021
							F1905	Device Revision or Replacement	A0705	Battery Problem	03/03/2021	03/03/2021
									A070504	Premature Discharge of Battery	03/03/2021	03/03/2021
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem	03/03/2021	03/03/2021
									A070504	Premature Discharge of Battery	03/03/2021	03/03/2021
							F08	Hospitalization or Prolonged Hospitalization	A0705	Battery Problem	03/03/2021	03/03/2021
									A070504	Premature Discharge of Battery	03/03/2021	03/03/2021
							F19	Surgical Intervention	A0705	Battery Problem	03/03/2021	03/03/2021
									A070504	Premature Discharge of Battery	03/03/2021	03/03/2021
							F1905	Device Revision or Replacement	A0705	Battery Problem	03/03/2021	03/03/2021
									A070504	Premature Discharge of Battery	03/03/2021	03/03/2021
					CT2727	Active implant devices/ancillaries	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem	03/03/2021	03/03/2021
									A070504	Premature Discharge of Battery	03/03/2021	03/03/2021
							F08	Hospitalization or Prolonged Hospitalization	A0705	Battery Problem	03/03/2021	03/03/2021
									A070504	Premature Discharge of Battery	03/03/2021	03/03/2021
							F19	Surgical Intervention	A0705	Battery Problem	03/03/2021	03/03/2021
									A070504	Premature Discharge of Battery	03/03/2021	03/03/2021
							F1905	Device Revision or Replacement	A0705	Battery Problem	03/03/2021	03/03/2021
									A070504	Premature Discharge of Battery	03/03/2021	03/03/2021
					CT2750	Implantable cardiac electrical stimulators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem	03/03/2021	03/03/2021
									A070504	Premature Discharge of Battery	03/03/2021	03/03/2021
							F08	Hospitalization or Prolonged Hospitalization	A0705	Battery Problem	03/03/2021	03/03/2021
									A070504	Premature Discharge of Battery	03/03/2021	03/03/2021
							F19	Surgical Intervention	A0705	Battery Problem	03/03/2021	03/03/2021
									A070504	Premature Discharge of Battery	03/03/2021	03/03/2021
							F1905	Device Revision or Replacement	A0705	Battery Problem	03/03/2021	03/03/2021
									A070504	Premature Discharge of Battery	03/03/2021	03/03/2021
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem	03/03/2021	03/03/2021
									A070504	Premature Discharge of Battery	03/03/2021	03/03/2021
							F08	Hospitalization or Prolonged Hospitalization	A0705	Battery Problem	03/03/2021	03/03/2021
									A070504	Premature Discharge of Battery	03/03/2021	03/03/2021
							F19	Surgical Intervention	A0705	Battery Problem	03/03/2021	03/03/2021
									A070504	Premature Discharge of Battery	03/03/2021	03/03/2021
							F1905	Device Revision or Replacement	A0705	Battery Problem	03/03/2021	03/03/2021
									A070504	Premature Discharge of Battery	03/03/2021	03/03/2021
					CT2985	Active	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem	03/03/2021	03/03/2021
									A070504	Premature Discharge of Battery	03/03/2021	03/03/2021
							F08	Hospitalization or Prolonged Hospitalization	A0705	Battery Problem	03/03/2021	03/03/2021
									A070504	Premature Discharge of Battery	03/03/2021	03/03/2021
							F19	Surgical Intervention	A0705	Battery Problem	03/03/2021	03/03/2021
									A070504	Premature Discharge of Battery	03/03/2021	03/03/2021
							F1905	Device Revision or Replacement	A0705	Battery Problem	03/03/2021	03/03/2021
									A070504	Premature Discharge of Battery	03/03/2021	03/03/2021
					CT3001	Implantable metal	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem	03/03/2021	03/03/2021
									A070504	Premature Discharge of Battery	03/03/2021	03/03/2021
							F08	Hospitalization or Prolonged Hospitalization	A0705	Battery Problem	03/03/2021	03/03/2021
									A070504	Premature Discharge of Battery	03/03/2021	03/03/2021
							F19	Surgical Intervention	A0705	Battery Problem	03/03/2021	03/03/2021
									A070504	Premature Discharge of Battery	03/03/2021	03/03/2021
							F1905	Device Revision or Replacement	A0705	Battery Problem	03/03/2021	03/03/2021
									A070504	Premature Discharge of Battery	03/03/2021	03/03/2021
					CT3002	Implantable synthetic polymer	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem	03/03/2021	03/03/2021
									A070504	Premature Discharge of Battery	03/03/2021	03/03/2021
							F08	Hospitalization or Prolonged Hospitalization	A0705	Battery Problem	03/03/2021	03/03/2021

									A070504	Premature Discharge of Battery	03/03/2021	03/03/2021
							F19	Surgical Intervention	A0705	Battery Problem	03/03/2021	03/03/2021
									A070504	Premature Discharge of Battery	03/03/2021	03/03/2021
							F1905	Device Revision or Replacement	A0705	Battery Problem	03/03/2021	03/03/2021
									A070504	Premature Discharge of Battery	03/03/2021	03/03/2021
					CT3006	Implantable inorganic material	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem	03/03/2021	03/03/2021
									A070504	Premature Discharge of Battery	03/03/2021	03/03/2021
							F08	Hospitalization or Prolonged Hospitalization	A0705	Battery Problem	03/03/2021	03/03/2021
									A070504	Premature Discharge of Battery	03/03/2021	03/03/2021
							F19	Surgical Intervention	A0705	Battery Problem	03/03/2021	03/03/2021
									A070504	Premature Discharge of Battery	03/03/2021	03/03/2021
							F1905	Device Revision or Replacement	A0705	Battery Problem	03/03/2021	03/03/2021
									A070504	Premature Discharge of Battery	03/03/2021	03/03/2021
					CT314	Electrically powered	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem	03/03/2021	03/03/2021
									A070504	Premature Discharge of Battery	03/03/2021	03/03/2021
							F08	Hospitalization or Prolonged Hospitalization	A0705	Battery Problem	03/03/2021	03/03/2021
									A070504	Premature Discharge of Battery	03/03/2021	03/03/2021
							F19	Surgical Intervention	A0705	Battery Problem	03/03/2021	03/03/2021
									A070504	Premature Discharge of Battery	03/03/2021	03/03/2021
							F1905	Device Revision or Replacement	A0705	Battery Problem	03/03/2021	03/03/2021
									A070504	Premature Discharge of Battery	03/03/2021	03/03/2021
					CT321	Long-term surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem	03/03/2021	03/03/2021
									A070504	Premature Discharge of Battery	03/03/2021	03/03/2021
							F08	Hospitalization or Prolonged Hospitalization	A0705	Battery Problem	03/03/2021	03/03/2021
									A070504	Premature Discharge of Battery	03/03/2021	03/03/2021
							F19	Surgical Intervention	A0705	Battery Problem	03/03/2021	03/03/2021
									A070504	Premature Discharge of Battery	03/03/2021	03/03/2021
							F1905	Device Revision or Replacement	A0705	Battery Problem	03/03/2021	03/03/2021
									A070504	Premature Discharge of Battery	03/03/2021	03/03/2021
					CT449	Generators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem	03/03/2021	03/03/2021
									A070504	Premature Discharge of Battery	03/03/2021	03/03/2021
							F08	Hospitalization or Prolonged Hospitalization	A0705	Battery Problem	03/03/2021	03/03/2021
									A070504	Premature Discharge of Battery	03/03/2021	03/03/2021
							F19	Surgical Intervention	A0705	Battery Problem	03/03/2021	03/03/2021
									A070504	Premature Discharge of Battery	03/03/2021	03/03/2021
							F1905	Device Revision or Replacement	A0705	Battery Problem	03/03/2021	03/03/2021
									A070504	Premature Discharge of Battery	03/03/2021	03/03/2021
					CT981	Single use	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem	03/03/2021	03/03/2021
									A070504	Premature Discharge of Battery	03/03/2021	03/03/2021
							F08	Hospitalization or Prolonged Hospitalization	A0705	Battery Problem	03/03/2021	03/03/2021
									A070504	Premature Discharge of Battery	03/03/2021	03/03/2021
							F19	Surgical Intervention	A0705	Battery Problem	03/03/2021	03/03/2021
									A070504	Premature Discharge of Battery	03/03/2021	03/03/2021
							F1905	Device Revision or Replacement	A0705	Battery Problem	03/03/2021	03/03/2021
									A070504	Premature Discharge of Battery	03/03/2021	03/03/2021
					CT983	Surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem	03/03/2021	03/03/2021
									A070504	Premature Discharge of Battery	03/03/2021	03/03/2021
							F08	Hospitalization or Prolonged Hospitalization	A0705	Battery Problem	03/03/2021	03/03/2021
									A070504	Premature Discharge of Battery	03/03/2021	03/03/2021
							F19	Surgical Intervention	A0705	Battery Problem	03/03/2021	03/03/2021
									A070504	Premature Discharge of Battery	03/03/2021	03/03/2021
							F1905	Device Revision or Replacement	A0705	Battery Problem	03/03/2021	03/03/2021
									A070504	Premature Discharge of Battery	03/03/2021	03/03/2021
10484	2021	All other reportable incidents	36241	Pacemaker/defibrillator lead	CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070909	Oversensing		
									A090801	Signal Artifact/Noise		
									A1301	Failure to Read Input Signal		
							F26	No Health Consequences or Impact	A070909	Oversensing		
									A090801	Signal Artifact/Noise		
									A1301	Failure to Read Input Signal		
10485	2021	All other reportable incidents	35852	Single-chamber implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	26/05/2021	26/05/2021
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	26/05/2021	26/05/2021
							F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A070504	Premature Discharge of Battery	26/05/2021	26/05/2021
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	26/05/2021	26/05/2021
10486	2021	All other reportable incidents	47265	Dual-chamber implantable pacemaker, rate-responsive	CT141	Electrodes/Leads and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	23/03/2021	23/03/2021
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	23/03/2021	23/03/2021
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	23/03/2021	23/03/2021
					CT145	Cardiovascular devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	23/03/2021	23/03/2021
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	23/03/2021	23/03/2021
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	23/03/2021	23/03/2021
					CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	23/03/2021	23/03/2021
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	23/03/2021	23/03/2021
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	23/03/2021	23/03/2021
					CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	23/03/2021	23/03/2021
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	23/03/2021	23/03/2021
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	23/03/2021	23/03/2021
					CT233	Surgical	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	23/03/2021	23/03/2021
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	23/03/2021	23/03/2021
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	23/03/2021	23/03/2021
					CT2406	Implantable	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	23/03/2021	23/03/2021
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	23/03/2021	23/03/2021
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	23/03/2021	23/03/2021
					CT2407	Active implantable	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	23/03/2021	23/03/2021
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	23/03/2021	23/03/2021
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	23/03/2021	23/03/2021
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	23/03/2021	23/03/2021
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	23/03/2021	23/03/2021
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	23/03/2021	23/03/2021
					CT2727	Active implant devices/ancillaries	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	23/03/2021	23/03/2021

							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	23/03/2021	23/03/2021
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	23/03/2021	23/03/2021
					CT2750	Implantable cardiac electrical stimulators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	23/03/2021	23/03/2021
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	23/03/2021	23/03/2021
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	23/03/2021	23/03/2021
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	23/03/2021	23/03/2021
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	23/03/2021	23/03/2021
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	23/03/2021	23/03/2021
					CT2985	Active	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	23/03/2021	23/03/2021
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	23/03/2021	23/03/2021
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	23/03/2021	23/03/2021
					CT3001	Implantable metal	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	23/03/2021	23/03/2021
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	23/03/2021	23/03/2021
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	23/03/2021	23/03/2021
					CT3002	Implantable synthetic polymer	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	23/03/2021	23/03/2021
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	23/03/2021	23/03/2021
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	23/03/2021	23/03/2021
					CT3006	Implantable inorganic material	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	23/03/2021	23/03/2021
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	23/03/2021	23/03/2021
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	23/03/2021	23/03/2021
					CT314	Electrically powered	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	23/03/2021	23/03/2021
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	23/03/2021	23/03/2021
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	23/03/2021	23/03/2021
					CT321	Long-term surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	23/03/2021	23/03/2021
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	23/03/2021	23/03/2021
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	23/03/2021	23/03/2021
					CT449	Generators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	23/03/2021	23/03/2021
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	23/03/2021	23/03/2021
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	23/03/2021	23/03/2021
					CT981	Single use	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	23/03/2021	23/03/2021
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	23/03/2021	23/03/2021
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	23/03/2021	23/03/2021
					CT983	Surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	23/03/2021	23/03/2021
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	23/03/2021	23/03/2021
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	23/03/2021	23/03/2021
10487	2021	All other reportable incidents	35852	Single-chamber implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	12/01/2021	12/01/2021
							F19	Surgical Intervention	A070504	Premature Discharge of Battery	12/01/2021	12/01/2021
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	12/01/2021	12/01/2021
10488	2021	All other reportable incidents	35852	Single-chamber implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem	26/01/2021	26/01/2021
									A070504	Premature Discharge of Battery	26/01/2021	26/01/2021
							F19	Surgical Intervention	A0705	Battery Problem	26/01/2021	26/01/2021
									A070504	Premature Discharge of Battery	26/01/2021	26/01/2021
10489	2021	All other reportable incidents	47265	Dual-chamber implantable pacemaker, rate-responsive	CT141	Electrodes/Leads and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	19/01/2021	19/01/2021
							F1901	Additional Surgery	A070504	Premature Discharge of Battery	19/01/2021	19/01/2021
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	19/01/2021	19/01/2021
					CT145	Cardiovascular devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	19/01/2021	19/01/2021
							F1901	Additional Surgery	A070504	Premature Discharge of Battery	19/01/2021	19/01/2021
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	19/01/2021	19/01/2021
					CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	19/01/2021	19/01/2021
							F1901	Additional Surgery	A070504	Premature Discharge of Battery	19/01/2021	19/01/2021
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	19/01/2021	19/01/2021
					CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	19/01/2021	19/01/2021
							F1901	Additional Surgery	A070504	Premature Discharge of Battery	19/01/2021	19/01/2021
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	19/01/2021	19/01/2021
					CT233	Surgical	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	19/01/2021	19/01/2021
							F1901	Additional Surgery	A070504	Premature Discharge of Battery	19/01/2021	19/01/2021
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	19/01/2021	19/01/2021
					CT2406	Implantable	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	19/01/2021	19/01/2021
							F1901	Additional Surgery	A070504	Premature Discharge of Battery	19/01/2021	19/01/2021
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	19/01/2021	19/01/2021
					CT2407	Active implantable	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	19/01/2021	19/01/2021
							F1901	Additional Surgery	A070504	Premature Discharge of Battery	19/01/2021	19/01/2021
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	19/01/2021	19/01/2021
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	19/01/2021	19/01/2021
							F1901	Additional Surgery	A070504	Premature Discharge of Battery	19/01/2021	19/01/2021
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	19/01/2021	19/01/2021
					CT2727	Active implant devices/ancillaries	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	19/01/2021	19/01/2021
							F1901	Additional Surgery	A070504	Premature Discharge of Battery	19/01/2021	19/01/2021
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	19/01/2021	19/01/2021
					CT2750	Implantable cardiac electrical stimulators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	19/01/2021	19/01/2021
							F1901	Additional Surgery	A070504	Premature Discharge of Battery	19/01/2021	19/01/2021
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	19/01/2021	19/01/2021
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	19/01/2021	19/01/2021
							F1901	Additional Surgery	A070504	Premature Discharge of Battery	19/01/2021	19/01/2021
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	19/01/2021	19/01/2021
					CT2985	Active	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	19/01/2021	19/01/2021
							F1901	Additional Surgery	A070504	Premature Discharge of Battery	19/01/2021	19/01/2021
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	19/01/2021	19/01/2021
					CT3001	Implantable metal	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	19/01/2021	19/01/2021
							F1901	Additional Surgery	A070504	Premature Discharge of Battery	19/01/2021	19/01/2021
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	19/01/2021	19/01/2021
					CT3002	Implantable synthetic polymer	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	19/01/2021	19/01/2021
							F1901	Additional Surgery	A070504	Premature Discharge of Battery	19/01/2021	19/01/2021
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	19/01/2021	19/01/2021
					CT3006	Implantable inorganic material	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	19/01/2021	19/01/2021
							F1901	Additional Surgery	A070504	Premature Discharge of Battery	19/01/2021	19/01/2021
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	19/01/2021	19/01/2021
					CT314	Electrically powered	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	19/01/2021	19/01/2021

							F1901	Additional Surgery	A070504	Premature Discharge of Battery	19/01/2021	19/01/2021
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	19/01/2021	19/01/2021
					CT321	Long-term surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	19/01/2021	19/01/2021
							F1901	Additional Surgery	A070504	Premature Discharge of Battery	19/01/2021	19/01/2021
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	19/01/2021	19/01/2021
					CT449	Generators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	19/01/2021	19/01/2021
							F1901	Additional Surgery	A070504	Premature Discharge of Battery	19/01/2021	19/01/2021
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	19/01/2021	19/01/2021
					CT981	Single use	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	19/01/2021	19/01/2021
							F1901	Additional Surgery	A070504	Premature Discharge of Battery	19/01/2021	19/01/2021
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	19/01/2021	19/01/2021
					CT983	Surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	19/01/2021	19/01/2021
							F1901	Additional Surgery	A070504	Premature Discharge of Battery	19/01/2021	19/01/2021
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	19/01/2021	19/01/2021
10490	2021	All other reportable incidents	35223	Endocardial/interventricular septal pacing lead	CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070909	Oversensing		
							F26	No Health Consequences or Impact	A070909	Oversensing		
10491	2021	All other reportable incidents	37265	Dual-chamber implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A160601	Premature Elective Replacement Indicator		
							F1905	Device Revision or Replacement	A160601	Premature Elective Replacement Indicator		
10492	2021	All other reportable incidents	35039	Epicardial pacing lead	CT1684	Cardiac pacemakers and associated devices	E060106	Heart Block	A040101	Fracture		
									A070101	Failure to Capture		
									A072201	High Impedance		
							F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A040101	Fracture		
									A070101	Failure to Capture		
									A072201	High Impedance		
							F1905	Device Revision or Replacement	A040101	Fracture		
									A070101	Failure to Capture		
									A072201	High Impedance		
10493	2021	All other reportable incidents	35852	Single-chamber implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E2104	Electric Shock	A0709	Device Sensing Problem		
									A070909	Oversensing		
									A071302	Inappropriate/Inadequate Shock/Stimulation		
							F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A0709	Device Sensing Problem		
									A070909	Oversensing		
									A071302	Inappropriate/Inadequate Shock/Stimulation		
							F15	Recognized Device or Procedural Complication	A0709	Device Sensing Problem		
									A070909	Oversensing		
									A071302	Inappropriate/Inadequate Shock/Stimulation		
10494	2021	All other reportable incidents	47267	Single-chamber implantable pacemaker, rate-responsive	CT141	Electrodes/Leads and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	12/01/2021	12/01/2021
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	12/01/2021	12/01/2021
							F12	Serious Injury/ Illness/ Impairment	A070504	Premature Discharge of Battery	12/01/2021	12/01/2021
							F1903	Device Explantation	A070504	Premature Discharge of Battery	12/01/2021	12/01/2021
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	12/01/2021	12/01/2021
					CT145	Cardiovascular devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	12/01/2021	12/01/2021
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	12/01/2021	12/01/2021
							F12	Serious Injury/ Illness/ Impairment	A070504	Premature Discharge of Battery	12/01/2021	12/01/2021
							F1903	Device Explantation	A070504	Premature Discharge of Battery	12/01/2021	12/01/2021
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	12/01/2021	12/01/2021
					CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	12/01/2021	12/01/2021
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	12/01/2021	12/01/2021
							F12	Serious Injury/ Illness/ Impairment	A070504	Premature Discharge of Battery	12/01/2021	12/01/2021
							F1903	Device Explantation	A070504	Premature Discharge of Battery	12/01/2021	12/01/2021
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	12/01/2021	12/01/2021
					CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	12/01/2021	12/01/2021
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	12/01/2021	12/01/2021
							F12	Serious Injury/ Illness/ Impairment	A070504	Premature Discharge of Battery	12/01/2021	12/01/2021
							F1903	Device Explantation	A070504	Premature Discharge of Battery	12/01/2021	12/01/2021
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	12/01/2021	12/01/2021
					CT233	Surgical	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	12/01/2021	12/01/2021
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	12/01/2021	12/01/2021
							F12	Serious Injury/ Illness/ Impairment	A070504	Premature Discharge of Battery	12/01/2021	12/01/2021
							F1903	Device Explantation	A070504	Premature Discharge of Battery	12/01/2021	12/01/2021
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	12/01/2021	12/01/2021
					CT2406	Implantable	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	12/01/2021	12/01/2021
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	12/01/2021	12/01/2021
							F12	Serious Injury/ Illness/ Impairment	A070504	Premature Discharge of Battery	12/01/2021	12/01/2021
							F1903	Device Explantation	A070504	Premature Discharge of Battery	12/01/2021	12/01/2021
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	12/01/2021	12/01/2021
					CT2407	Active implantable	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	12/01/2021	12/01/2021
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	12/01/2021	12/01/2021
							F12	Serious Injury/ Illness/ Impairment	A070504	Premature Discharge of Battery	12/01/2021	12/01/2021
							F1903	Device Explantation	A070504	Premature Discharge of Battery	12/01/2021	12/01/2021
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	12/01/2021	12/01/2021
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	12/01/2021	12/01/2021
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	12/01/2021	12/01/2021
							F12	Serious Injury/ Illness/ Impairment	A070504	Premature Discharge of Battery	12/01/2021	12/01/2021
							F1903	Device Explantation	A070504	Premature Discharge of Battery	12/01/2021	12/01/2021
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	12/01/2021	12/01/2021
					CT2727	Active implant devices/ancillaries	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	12/01/2021	12/01/2021
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	12/01/2021	12/01/2021
							F12	Serious Injury/ Illness/ Impairment	A070504	Premature Discharge of Battery	12/01/2021	12/01/2021
							F1903	Device Explantation	A070504	Premature Discharge of Battery	12/01/2021	12/01/2021
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	12/01/2021	12/01/2021
					CT2750	Implantable cardiac electrical stimulators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	12/01/2021	12/01/2021
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	12/01/2021	12/01/2021
							F12	Serious Injury/ Illness/ Impairment	A070504	Premature Discharge of Battery	12/01/2021	12/01/2021
							F1903	Device Explantation	A070504	Premature Discharge of Battery	12/01/2021	12/01/2021
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	12/01/2021	12/01/2021
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	12/01/2021	12/01/2021
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	12/01/2021	12/01/2021

							F12	Serious Injury/ Illness/ Impairment	A070504	Premature Discharge of Battery	12/01/2021	12/01/2021
							F1903	Device Explantation	A070504	Premature Discharge of Battery	12/01/2021	12/01/2021
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	12/01/2021	12/01/2021
					CT2985	Active	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	12/01/2021	12/01/2021
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	12/01/2021	12/01/2021
							F12	Serious Injury/ Illness/ Impairment	A070504	Premature Discharge of Battery	12/01/2021	12/01/2021
							F1903	Device Explantation	A070504	Premature Discharge of Battery	12/01/2021	12/01/2021
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	12/01/2021	12/01/2021
					CT3001	Implantable metal	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	12/01/2021	12/01/2021
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	12/01/2021	12/01/2021
							F12	Serious Injury/ Illness/ Impairment	A070504	Premature Discharge of Battery	12/01/2021	12/01/2021
							F1903	Device Explantation	A070504	Premature Discharge of Battery	12/01/2021	12/01/2021
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	12/01/2021	12/01/2021
					CT3002	Implantable synthetic polymer	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	12/01/2021	12/01/2021
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	12/01/2021	12/01/2021
							F12	Serious Injury/ Illness/ Impairment	A070504	Premature Discharge of Battery	12/01/2021	12/01/2021
							F1903	Device Explantation	A070504	Premature Discharge of Battery	12/01/2021	12/01/2021
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	12/01/2021	12/01/2021
					CT3006	Implantable inorganic material	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	12/01/2021	12/01/2021
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	12/01/2021	12/01/2021
							F12	Serious Injury/ Illness/ Impairment	A070504	Premature Discharge of Battery	12/01/2021	12/01/2021
							F1903	Device Explantation	A070504	Premature Discharge of Battery	12/01/2021	12/01/2021
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	12/01/2021	12/01/2021
					CT314	Electrically powered	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	12/01/2021	12/01/2021
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	12/01/2021	12/01/2021
							F12	Serious Injury/ Illness/ Impairment	A070504	Premature Discharge of Battery	12/01/2021	12/01/2021
							F1903	Device Explantation	A070504	Premature Discharge of Battery	12/01/2021	12/01/2021
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	12/01/2021	12/01/2021
					CT321	Long-term surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	12/01/2021	12/01/2021
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	12/01/2021	12/01/2021
							F12	Serious Injury/ Illness/ Impairment	A070504	Premature Discharge of Battery	12/01/2021	12/01/2021
							F1903	Device Explantation	A070504	Premature Discharge of Battery	12/01/2021	12/01/2021
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	12/01/2021	12/01/2021
					CT449	Generators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	12/01/2021	12/01/2021
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	12/01/2021	12/01/2021
							F12	Serious Injury/ Illness/ Impairment	A070504	Premature Discharge of Battery	12/01/2021	12/01/2021
							F1903	Device Explantation	A070504	Premature Discharge of Battery	12/01/2021	12/01/2021
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	12/01/2021	12/01/2021
					CT981	Single use	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	12/01/2021	12/01/2021
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	12/01/2021	12/01/2021
							F12	Serious Injury/ Illness/ Impairment	A070504	Premature Discharge of Battery	12/01/2021	12/01/2021
							F1903	Device Explantation	A070504	Premature Discharge of Battery	12/01/2021	12/01/2021
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	12/01/2021	12/01/2021
					CT983	Surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	12/01/2021	12/01/2021
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	12/01/2021	12/01/2021
							F12	Serious Injury/ Illness/ Impairment	A070504	Premature Discharge of Battery	12/01/2021	12/01/2021
							F1903	Device Explantation	A070504	Premature Discharge of Battery	12/01/2021	12/01/2021
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	12/01/2021	12/01/2021
10495	2021	All other reportable incidents	47270	Cardiac resynchronization therapy implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A070101	Failure to Capture		
									A070909	Oversensing		
									A090801	Signal Artifact/Noise		
									A1301	Failure to Read Input Signal		
							F26	No Health Consequences or Impact	A070101	Failure to Capture		
									A070909	Oversensing		
									A090801	Signal Artifact/Noise		
									A1301	Failure to Read Input Signal		
10496	2021	All other reportable incidents	35223	Endocardial/interventricular septal pacing lead	CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A040101	Fracture		
									A070909	Oversensing		
									A090801	Signal Artifact/Noise		
							F26	No Health Consequences or Impact	A040101	Fracture		
									A070909	Oversensing		
									A090801	Signal Artifact/Noise		
10497	2021	All other reportable incidents	35852	Single-chamber implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem	08/01/2021	08/01/2021
									A070504	Premature Discharge of Battery	08/01/2021	08/01/2021
							F19	Surgical Intervention	A0705	Battery Problem	08/01/2021	08/01/2021
									A070504	Premature Discharge of Battery	08/01/2021	08/01/2021
							F1905	Device Revision or Replacement	A0705	Battery Problem	08/01/2021	08/01/2021
									A070504	Premature Discharge of Battery	08/01/2021	08/01/2021
10498	2021	All other reportable incidents	35852	Single-chamber implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem	26/01/2021	26/01/2021
							F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A0705	Battery Problem	26/01/2021	26/01/2021
							F19	Surgical Intervention	A0705	Battery Problem	26/01/2021	26/01/2021
10499	2021	All other reportable incidents	47270	Cardiac resynchronization therapy implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A070909	Oversensing		
							F26	No Health Consequences or Impact	A070909	Oversensing		
10500	2021	All other reportable incidents	36241	Pacemaker/defibrillator lead	CT1684	Cardiac pacemakers and associated devices	E2104	Electric Shock	A04	Material Integrity Problem	22/12/2020	22/12/2020
							F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A04	Material Integrity Problem	22/12/2020	22/12/2020
							F1905	Device Revision or Replacement	A04	Material Integrity Problem	22/12/2020	22/12/2020
10501	2021	All other reportable incidents	47270	Cardiac resynchronization therapy implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A072201	High Impedance		
							F26	No Health Consequences or Impact	A072201	High Impedance		
10502	2021	All other reportable incidents	47265	Dual-chamber implantable pacemaker, rate-responsive	CT141	Electrodes/Leads and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT145	Cardiovascular devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT233	Surgical	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		

					CT2406	Implantable	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT2407	Active implantable	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT2727	Active implant devices/ancillaries	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT2750	Implantable cardiac electrical stimulators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT2985	Active	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT3001	Implantable metal	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT3002	Implantable synthetic polymer	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT3006	Implantable inorganic material	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT314	Electrically powered	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT321	Long-term surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT449	Generators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT981	Single use	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT983	Surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
10503	2021	All other reportable incidents	35223	Endocardial/interventricular septal pacing lead	CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A0401	Break		
									A070102	High Capture Threshold		
									A070103	Intermittent Capture		
									A070901	Decreased Sensitivity		
									A070909	Oversensing		
									A0722	Impedance Problem		
							F12	Serious Injury/ Illness/ Impairment	A0401	Break		
									A070102	High Capture Threshold		
									A070103	Intermittent Capture		
									A070901	Decreased Sensitivity		
									A070909	Oversensing		
									A0722	Impedance Problem		
							F23	Unexpected Medical Intervention	A0401	Break		
									A070102	High Capture Threshold		
									A070103	Intermittent Capture		
									A070901	Decreased Sensitivity		
									A070909	Oversensing		
									A0722	Impedance Problem		
							F26	No Health Consequences or Impact	A0401	Break		
									A070102	High Capture Threshold		
									A070103	Intermittent Capture		
									A070901	Decreased Sensitivity		
									A070909	Oversensing		
									A0722	Impedance Problem		
10504	2021	All other reportable incidents	36241	Pacemaker/defibrillator lead	CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070909	Oversensing		
									A1301	Failure to Read Input Signal		
							F15	Recognized Device or Procedural Complication	A070909	Oversensing		
									A1301	Failure to Read Input Signal		
10505	2021	All other reportable incidents	17882	Physiologic-monitoring defibrillation system	CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070503	Failure to Run on Battery		
							F27	Problem Identified During Non-Clinical Procedure	A070503	Failure to Run on Battery		
10506	2021	All other reportable incidents	47265	Dual-chamber implantable pacemaker, rate-responsive	CT141	Electrodes/Leads and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem		
									A0722	Impedance Problem		
							F26	No Health Consequences or Impact	A0705	Battery Problem		
									A0722	Impedance Problem		
					CT145	Cardiovascular devices	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem		
									A0722	Impedance Problem		
							F26	No Health Consequences or Impact	A0705	Battery Problem		
									A0722	Impedance Problem		
					CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem		
									A0722	Impedance Problem		
							F26	No Health Consequences or Impact	A0705	Battery Problem		
									A0722	Impedance Problem		
					CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem		
									A0722	Impedance Problem		
							F26	No Health Consequences or Impact	A0705	Battery Problem		
									A0722	Impedance Problem		
					CT233	Surgical	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem		
									A0722	Impedance Problem		
							F26	No Health Consequences or Impact	A0705	Battery Problem		
									A0722	Impedance Problem		
					CT2406	Implantable	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem		
									A0722	Impedance Problem		
							F26	No Health Consequences or Impact	A0705	Battery Problem		
									A0722	Impedance Problem		
					CT2407	Active implantable	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem		
									A0722	Impedance Problem		
							F26	No Health Consequences or Impact	A0705	Battery Problem		
									A0722	Impedance Problem		

					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem		
									A0722	Impedance Problem		
							F26	No Health Consequences or Impact	A0705	Battery Problem		
									A0722	Impedance Problem		
					CT2727	Active implant devices/ancillaries	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem		
									A0722	Impedance Problem		
							F26	No Health Consequences or Impact	A0705	Battery Problem		
									A0722	Impedance Problem		
					CT2750	Implantable cardiac electrical stimulators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem		
									A0722	Impedance Problem		
							F26	No Health Consequences or Impact	A0705	Battery Problem		
									A0722	Impedance Problem		
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem		
									A0722	Impedance Problem		
							F26	No Health Consequences or Impact	A0705	Battery Problem		
									A0722	Impedance Problem		
					CT2985	Active	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem		
									A0722	Impedance Problem		
							F26	No Health Consequences or Impact	A0705	Battery Problem		
									A0722	Impedance Problem		
					CT3001	Implantable metal	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem		
									A0722	Impedance Problem		
							F26	No Health Consequences or Impact	A0705	Battery Problem		
									A0722	Impedance Problem		
					CT3002	Implantable synthetic polymer	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem		
									A0722	Impedance Problem		
							F26	No Health Consequences or Impact	A0705	Battery Problem		
									A0722	Impedance Problem		
					CT3006	Implantable inorganic material	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem		
									A0722	Impedance Problem		
							F26	No Health Consequences or Impact	A0705	Battery Problem		
									A0722	Impedance Problem		
					CT314	Electrically powered	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem		
									A0722	Impedance Problem		
							F26	No Health Consequences or Impact	A0705	Battery Problem		
									A0722	Impedance Problem		
					CT321	Long-term surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem		
									A0722	Impedance Problem		
							F26	No Health Consequences or Impact	A0705	Battery Problem		
									A0722	Impedance Problem		
					CT449	Generators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem		
									A0722	Impedance Problem		
							F26	No Health Consequences or Impact	A0705	Battery Problem		
									A0722	Impedance Problem		
					CT981	Single use	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem		
									A0722	Impedance Problem		
							F26	No Health Consequences or Impact	A0705	Battery Problem		
									A0722	Impedance Problem		
					CT983	Surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem		
									A0722	Impedance Problem		
							F26	No Health Consequences or Impact	A0705	Battery Problem		
									A0722	Impedance Problem		
10507	2021	All other reportable incidents	47265	Dual-chamber implantable pacemaker, rate-responsive	CT141	Electrodes/Leads and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A0722	Impedance Problem		
							F26	No Health Consequences or Impact	A0722	Impedance Problem		
					CT145	Cardiovascular devices	E2403	No Clinical Signs, Symptoms or Conditions	A0722	Impedance Problem		
							F26	No Health Consequences or Impact	A0722	Impedance Problem		
					CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A0722	Impedance Problem		
							F26	No Health Consequences or Impact	A0722	Impedance Problem		
					CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A0722	Impedance Problem		
							F26	No Health Consequences or Impact	A0722	Impedance Problem		
					CT233	Surgical	E2403	No Clinical Signs, Symptoms or Conditions	A0722	Impedance Problem		
							F26	No Health Consequences or Impact	A0722	Impedance Problem		
					CT2406	Implantable	E2403	No Clinical Signs, Symptoms or Conditions	A0722	Impedance Problem		
							F26	No Health Consequences or Impact	A0722	Impedance Problem		
					CT2407	Active implantable	E2403	No Clinical Signs, Symptoms or Conditions	A0722	Impedance Problem		
							F26	No Health Consequences or Impact	A0722	Impedance Problem		
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A0722	Impedance Problem		
							F26	No Health Consequences or Impact	A0722	Impedance Problem		
					CT2727	Active implant devices/ancillaries	E2403	No Clinical Signs, Symptoms or Conditions	A0722	Impedance Problem		
							F26	No Health Consequences or Impact	A0722	Impedance Problem		
					CT2750	Implantable cardiac electrical stimulators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A0722	Impedance Problem		
							F26	No Health Consequences or Impact	A0722	Impedance Problem		
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A0722	Impedance Problem		
							F26	No Health Consequences or Impact	A0722	Impedance Problem		
					CT2985	Active	E2403	No Clinical Signs, Symptoms or Conditions	A0722	Impedance Problem		
							F26	No Health Consequences or Impact	A0722	Impedance Problem		
					CT3001	Implantable metal	E2403	No Clinical Signs, Symptoms or Conditions	A0722	Impedance Problem		
							F26	No Health Consequences or Impact	A0722	Impedance Problem		
					CT3002	Implantable synthetic polymer	E2403	No Clinical Signs, Symptoms or Conditions	A0722	Impedance Problem		
							F26	No Health Consequences or Impact	A0722	Impedance Problem		
					CT3006	Implantable inorganic material	E2403	No Clinical Signs, Symptoms or Conditions	A0722	Impedance Problem		
							F26	No Health Consequences or Impact	A0722	Impedance Problem		
					CT314	Electrically powered	E2403	No Clinical Signs, Symptoms or Conditions	A0722	Impedance Problem		
							F26	No Health Consequences or Impact	A0722	Impedance Problem		
					CT321	Long-term surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A0722	Impedance Problem		
							F26	No Health Consequences or Impact	A0722	Impedance Problem		
					CT449	Generators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A0722	Impedance Problem		
							F26	No Health Consequences or Impact	A0722	Impedance Problem		

					CT981	Single use	E2403	No Clinical Signs, Symptoms or Conditions	A0722	Impedance Problem		
							F26	No Health Consequences or Impact	A0722	Impedance Problem		
					CT983	Surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A0722	Impedance Problem		
							F26	No Health Consequences or Impact	A0722	Impedance Problem		
10508	2021	All other reportable incidents	17882	Physiologic-monitoring defibrillation system	CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A0719	Unexpected Shutdown		
							F26	No Health Consequences or Impact	A0719	Unexpected Shutdown		
10509	2021	All other reportable incidents	35223	Endocardial/interventricular septal pacing lead	CT1684	Cardiac pacemakers and associated devices	E2401	Insufficient Information	A07	Electrical /Electronic Property Problem		
									A070909	Oversensing		
									A072202	Low Impedance		
							E2403	No Clinical Signs, Symptoms or Conditions	A07	Electrical /Electronic Property Problem		
									A070909	Oversensing		
									A072202	Low Impedance		
							F23	Unexpected Medical Intervention	A07	Electrical /Electronic Property Problem		
									A070909	Oversensing		
									A072202	Low Impedance		
10510	2021	All other reportable incidents	36241	Pacemaker/defibrillator lead	CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070101	Failure to Capture	23/12/2020	23/12/2020
							F08	Hospitalization or Prolonged Hospitalization	A070101	Failure to Capture	23/12/2020	23/12/2020
							F12	Serious Injury/ Illness/ Impairment	A070101	Failure to Capture	23/12/2020	23/12/2020
							F1905	Device Revision or Replacement	A070101	Failure to Capture	23/12/2020	23/12/2020
10511	2021	All other reportable incidents	36241	Pacemaker/defibrillator lead	CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070101	Failure to Capture	23/12/2020	23/12/2020
							F08	Hospitalization or Prolonged Hospitalization	A070101	Failure to Capture	23/12/2020	23/12/2020
							F12	Serious Injury/ Illness/ Impairment	A070101	Failure to Capture	23/12/2020	23/12/2020
							F1905	Device Revision or Replacement	A070101	Failure to Capture	23/12/2020	23/12/2020
10512	2021	All other reportable incidents	35852	Single-chamber implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A072201	High Impedance		
							F15	Recognized Device or Procedural Complication	A072201	High Impedance		
10513	2021	All other reportable incidents	47265	Dual-chamber implantable pacemaker, rate-responsive	CT141	Electrodes/Leads and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT145	Cardiovascular devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT233	Surgical	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT2406	Implantable	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT2407	Active implantable	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT2727	Active implant devices/ancillaries	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT2750	Implantable cardiac electrical stimulators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT2985	Active	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT3001	Implantable metal	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT3002	Implantable synthetic polymer	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT3006	Implantable inorganic material	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT314	Electrically powered	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT321	Long-term surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT449	Generators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT981	Single use	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT983	Surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
10514	2021	All other reportable incidents	35852	Single-chamber implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A070909	Oversensing		
									A071302	Inappropriate/Inadequate Shock/Stimulation		
									A090801	Signal Artifact/Noise		
							F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A070909	Oversensing		
									A071302	Inappropriate/Inadequate Shock/Stimulation		
									A090801	Signal Artifact/Noise		
							F26	No Health Consequences or Impact	A070909	Oversensing		
									A071302	Inappropriate/Inadequate Shock/Stimulation		
									A090801	Signal Artifact/Noise		
10515	2021	All other reportable incidents	35223	Endocardial/interventricular septal pacing lead	CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A040101	Fracture		
									A070909	Oversensing		
									A0722	Impedance Problem		
									A072201	High Impedance		
									A090801	Signal Artifact/Noise		
							F26	No Health Consequences or Impact	A040101	Fracture		
									A070909	Oversensing		
									A0722	Impedance Problem		
									A072201	High Impedance		
									A090801	Signal Artifact/Noise		
10516	2021	All other reportable incidents	35852	Single-chamber implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A070909	Oversensing		
									A071302	Inappropriate/Inadequate Shock/Stimulation		
									A1301	Failure to Read Input Signal		
							F26	No Health Consequences or Impact	A070909	Oversensing		

									A071302	Inappropriate/Inadequate Shock/Stimulation		
									A1301	Failure to Read Input Signal		
10517	2021	All other reportable incidents	36241	Pacemaker/defibrillator lead	CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070909	Oversensing		
							F26	No Health Consequences or Impact	A070909	Oversensing		
10518	2021	All other reportable incidents	35852	Single-chamber implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A1107	Data Problem		
							F26	No Health Consequences or Impact	A1107	Data Problem		
10519	2021	All other reportable incidents	47263	Cardiac resynchronization therapy implantable pacemaker	CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A0712	Pacing Problem		
									A23	Use of Device Problem		
							F26	No Health Consequences or Impact	A0712	Pacing Problem		
									A23	Use of Device Problem		
					CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A0712	Pacing Problem		
									A23	Use of Device Problem		
							F26	No Health Consequences or Impact	A0712	Pacing Problem		
									A23	Use of Device Problem		
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A0712	Pacing Problem		
									A23	Use of Device Problem		
							F26	No Health Consequences or Impact	A0712	Pacing Problem		
									A23	Use of Device Problem		
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A0712	Pacing Problem		
									A23	Use of Device Problem		
							F26	No Health Consequences or Impact	A0712	Pacing Problem		
									A23	Use of Device Problem		
10520	2021	All other reportable incidents	36241	Pacemaker/defibrillator lead	CT1684	Cardiac pacemakers and associated devices	E2104	Electric Shock	A072202	Low Impedance		
									A090801	Signal Artifact/Noise		
							F08	Hospitalization or Prolonged Hospitalization	A072202	Low Impedance		
									A090801	Signal Artifact/Noise		
							F1905	Device Revision or Replacement	A072202	Low Impedance		
									A090801	Signal Artifact/Noise		
10521	2021	All other reportable incidents	37265	Dual-chamber implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E2104	Electric Shock	A070602	Delayed Charge Time	13/01/2021	13/01/2021
									A072202	Low Impedance	13/01/2021	13/01/2021
									A0908	Incorrect, Inadequate or Imprecise Result or Readings	13/01/2021	13/01/2021
									A090801	Signal Artifact/Noise	13/01/2021	13/01/2021
							F08	Hospitalization or Prolonged Hospitalization	A070602	Delayed Charge Time	13/01/2021	13/01/2021
									A072202	Low Impedance	13/01/2021	13/01/2021
									A0908	Incorrect, Inadequate or Imprecise Result or Readings	13/01/2021	13/01/2021
									A090801	Signal Artifact/Noise	13/01/2021	13/01/2021
							F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A070602	Delayed Charge Time	13/01/2021	13/01/2021
									A072202	Low Impedance	13/01/2021	13/01/2021
									A0908	Incorrect, Inadequate or Imprecise Result or Readings	13/01/2021	13/01/2021
									A090801	Signal Artifact/Noise	13/01/2021	13/01/2021
							F1905	Device Revision or Replacement	A070602	Delayed Charge Time	13/01/2021	13/01/2021
									A072202	Low Impedance	13/01/2021	13/01/2021
									A0908	Incorrect, Inadequate or Imprecise Result or Readings	13/01/2021	13/01/2021
									A090801	Signal Artifact/Noise	13/01/2021	13/01/2021
10522	2021	All other reportable incidents	36241	Pacemaker/defibrillator lead	CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A051201	Device Dislodged or Dislocated	30/12/2020	30/12/2020
									A1502	Positioning Problem	30/12/2020	30/12/2020
							F08	Hospitalization or Prolonged Hospitalization	A051201	Device Dislodged or Dislocated	30/12/2020	30/12/2020
									A1502	Positioning Problem	30/12/2020	30/12/2020
							F14	Prolonged Episode of Care	A051201	Device Dislodged or Dislocated	30/12/2020	30/12/2020
									A1502	Positioning Problem	30/12/2020	30/12/2020
							F19	Surgical Intervention	A051201	Device Dislodged or Dislocated	30/12/2020	30/12/2020
									A1502	Positioning Problem	30/12/2020	30/12/2020
10523	2021	All other reportable incidents	35852	Single-chamber implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E2104	Electric Shock	A071302	Inappropriate/Inadequate Shock/Stimulation		
							F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A071302	Inappropriate/Inadequate Shock/Stimulation		
							F26	No Health Consequences or Impact	A071302	Inappropriate/Inadequate Shock/Stimulation		
10524	2021	All other reportable incidents	35852	Single-chamber implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E2104	Electric Shock	A071302	Inappropriate/Inadequate Shock/Stimulation	12/01/2021	12/01/2021
							F08	Hospitalization or Prolonged Hospitalization	A071302	Inappropriate/Inadequate Shock/Stimulation	12/01/2021	12/01/2021
							F1905	Device Revision or Replacement	A071302	Inappropriate/Inadequate Shock/Stimulation	12/01/2021	12/01/2021
10525	2021	All other reportable incidents	47263	Cardiac resynchronization therapy implantable pacemaker	CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A0712	Pacing Problem		
							F26	No Health Consequences or Impact	A0712	Pacing Problem		
					CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A0712	Pacing Problem		
							F26	No Health Consequences or Impact	A0712	Pacing Problem		
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A0712	Pacing Problem		
							F26	No Health Consequences or Impact	A0712	Pacing Problem		
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A0712	Pacing Problem		
							F26	No Health Consequences or Impact	A0712	Pacing Problem		
10526	2021	All other reportable incidents	47263	Cardiac resynchronization therapy implantable pacemaker	CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	13/01/2021	13/01/2021
							F1901	Additional Surgery	A070504	Premature Discharge of Battery	13/01/2021	13/01/2021
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	13/01/2021	13/01/2021
					CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	13/01/2021	13/01/2021
							F1901	Additional Surgery	A070504	Premature Discharge of Battery	13/01/2021	13/01/2021
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	13/01/2021	13/01/2021
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	13/01/2021	13/01/2021
							F1901	Additional Surgery	A070504	Premature Discharge of Battery	13/01/2021	13/01/2021
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	13/01/2021	13/01/2021
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	13/01/2021	13/01/2021
							F1901	Additional Surgery	A070504	Premature Discharge of Battery	13/01/2021	13/01/2021
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	13/01/2021	13/01/2021
10527	2021	All other reportable incidents	35852	Single-chamber implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem		
							F26	No Health Consequences or Impact	A0705	Battery Problem		
10528	2021	All other reportable incidents	47267	Single-chamber implantable pacemaker, rate-responsive	CT141	Electrodes/Leads and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
							F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A0705	Battery Problem		

								A070504	Premature Discharge of Battery		
				CT145	Cardiovascular devices	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem		
								A070504	Premature Discharge of Battery		
						F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A0705	Battery Problem		
								A070504	Premature Discharge of Battery		
						F26	No Health Consequences or Impact	A0705	Battery Problem		
								A070504	Premature Discharge of Battery		
				CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem		
								A070504	Premature Discharge of Battery		
						F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A0705	Battery Problem		
								A070504	Premature Discharge of Battery		
						F26	No Health Consequences or Impact	A0705	Battery Problem		
								A070504	Premature Discharge of Battery		
				CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem		
								A070504	Premature Discharge of Battery		
						F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A0705	Battery Problem		
								A070504	Premature Discharge of Battery		
						F26	No Health Consequences or Impact	A0705	Battery Problem		
								A070504	Premature Discharge of Battery		
				CT233	Surgical	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem		
								A070504	Premature Discharge of Battery		
						F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A0705	Battery Problem		
								A070504	Premature Discharge of Battery		
						F26	No Health Consequences or Impact	A0705	Battery Problem		
								A070504	Premature Discharge of Battery		
				CT2406	Implantable	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem		
								A070504	Premature Discharge of Battery		
						F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A0705	Battery Problem		
								A070504	Premature Discharge of Battery		
						F26	No Health Consequences or Impact	A0705	Battery Problem		
								A070504	Premature Discharge of Battery		
				CT2407	Active implantable	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem		
								A070504	Premature Discharge of Battery		
						F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A0705	Battery Problem		
								A070504	Premature Discharge of Battery		
						F26	No Health Consequences or Impact	A0705	Battery Problem		
								A070504	Premature Discharge of Battery		
				CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem		
								A070504	Premature Discharge of Battery		
						F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A0705	Battery Problem		
								A070504	Premature Discharge of Battery		
						F26	No Health Consequences or Impact	A0705	Battery Problem		
								A070504	Premature Discharge of Battery		
				CT2727	Active implant devices/ancillaries	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem		
								A070504	Premature Discharge of Battery		
						F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A0705	Battery Problem		
								A070504	Premature Discharge of Battery		
						F26	No Health Consequences or Impact	A0705	Battery Problem		
								A070504	Premature Discharge of Battery		
				CT2750	Implantable cardiac electrical stimulators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem		
								A070504	Premature Discharge of Battery		
						F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A0705	Battery Problem		
								A070504	Premature Discharge of Battery		
						F26	No Health Consequences or Impact	A0705	Battery Problem		
								A070504	Premature Discharge of Battery		
				CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem		
								A070504	Premature Discharge of Battery		
						F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A0705	Battery Problem		
								A070504	Premature Discharge of Battery		
						F26	No Health Consequences or Impact	A0705	Battery Problem		
								A070504	Premature Discharge of Battery		
				CT2985	Active	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem		
								A070504	Premature Discharge of Battery		
						F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A0705	Battery Problem		
								A070504	Premature Discharge of Battery		
						F26	No Health Consequences or Impact	A0705	Battery Problem		
								A070504	Premature Discharge of Battery		
				CT3001	Implantable metal	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem		
								A070504	Premature Discharge of Battery		
						F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A0705	Battery Problem		
								A070504	Premature Discharge of Battery		
						F26	No Health Consequences or Impact	A0705	Battery Problem		
								A070504	Premature Discharge of Battery		
				CT3002	Implantable synthetic polymer	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem		
								A070504	Premature Discharge of Battery		
						F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A0705	Battery Problem		
								A070504	Premature Discharge of Battery		
						F26	No Health Consequences or Impact	A0705	Battery Problem		
								A070504	Premature Discharge of Battery		
				CT3006	Implantable inorganic material	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem		
								A070504	Premature Discharge of Battery		
						F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A0705	Battery Problem		
								A070504	Premature Discharge of Battery		
						F26	No Health Consequences or Impact	A0705	Battery Problem		
								A070504	Premature Discharge of Battery		
				CT314	Electrically powered	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem		
								A070504	Premature Discharge of Battery		
						F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A0705	Battery Problem		

									A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
					CT321	Long-term surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
							F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
					CT449	Generators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
							F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
					CT981	Single use	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
							F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
					CT983	Surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
							F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
10529	2021	All other reportable incidents	35852	Single-chamber implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem	02/02/2021	02/02/2021
									A070504	Premature Discharge of Battery	02/02/2021	02/02/2021
							F19	Surgical Intervention	A0705	Battery Problem	02/02/2021	02/02/2021
									A070504	Premature Discharge of Battery	02/02/2021	02/02/2021
10530	2021	All other reportable incidents	35852	Single-chamber implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem	23/02/2021	23/02/2021
									A070504	Premature Discharge of Battery	23/02/2021	23/02/2021
							F19	Surgical Intervention	A0705	Battery Problem	23/02/2021	23/02/2021
									A070504	Premature Discharge of Battery	23/02/2021	23/02/2021
10531	2021	All other reportable incidents	47263	Cardiac resynchronization therapy implantable pacemaker	CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A090804	False Positive Result		
							F26	No Health Consequences or Impact	A090804	False Positive Result		
					CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A090804	False Positive Result		
							F26	No Health Consequences or Impact	A090804	False Positive Result		
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A090804	False Positive Result		
							F26	No Health Consequences or Impact	A090804	False Positive Result		
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A090804	False Positive Result		
							F26	No Health Consequences or Impact	A090804	False Positive Result		
10532	2021	All other reportable incidents	47265	Dual-chamber implantable pacemaker, rate-responsive	CT141	Electrodes/Leads and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	25/01/2021	25/01/2021
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	25/01/2021	25/01/2021
							F12	Serious Injury/ Illness/ Impairment	A070504	Premature Discharge of Battery	25/01/2021	25/01/2021
							F19	Surgical Intervention	A070504	Premature Discharge of Battery	25/01/2021	25/01/2021
					CT145	Cardiovascular devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	25/01/2021	25/01/2021
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	25/01/2021	25/01/2021
							F12	Serious Injury/ Illness/ Impairment	A070504	Premature Discharge of Battery	25/01/2021	25/01/2021
							F19	Surgical Intervention	A070504	Premature Discharge of Battery	25/01/2021	25/01/2021
					CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	25/01/2021	25/01/2021
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	25/01/2021	25/01/2021
							F12	Serious Injury/ Illness/ Impairment	A070504	Premature Discharge of Battery	25/01/2021	25/01/2021
							F19	Surgical Intervention	A070504	Premature Discharge of Battery	25/01/2021	25/01/2021
					CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	25/01/2021	25/01/2021
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	25/01/2021	25/01/2021
							F12	Serious Injury/ Illness/ Impairment	A070504	Premature Discharge of Battery	25/01/2021	25/01/2021
							F19	Surgical Intervention	A070504	Premature Discharge of Battery	25/01/2021	25/01/2021
					CT233	Surgical	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	25/01/2021	25/01/2021
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	25/01/2021	25/01/2021
							F12	Serious Injury/ Illness/ Impairment	A070504	Premature Discharge of Battery	25/01/2021	25/01/2021
							F19	Surgical Intervention	A070504	Premature Discharge of Battery	25/01/2021	25/01/2021
					CT2406	Implantable	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	25/01/2021	25/01/2021
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	25/01/2021	25/01/2021
							F12	Serious Injury/ Illness/ Impairment	A070504	Premature Discharge of Battery	25/01/2021	25/01/2021
							F19	Surgical Intervention	A070504	Premature Discharge of Battery	25/01/2021	25/01/2021
					CT2407	Active implantable	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	25/01/2021	25/01/2021
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	25/01/2021	25/01/2021
							F12	Serious Injury/ Illness/ Impairment	A070504	Premature Discharge of Battery	25/01/2021	25/01/2021
							F19	Surgical Intervention	A070504	Premature Discharge of Battery	25/01/2021	25/01/2021
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	25/01/2021	25/01/2021
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	25/01/2021	25/01/2021
							F12	Serious Injury/ Illness/ Impairment	A070504	Premature Discharge of Battery	25/01/2021	25/01/2021
							F19	Surgical Intervention	A070504	Premature Discharge of Battery	25/01/2021	25/01/2021
					CT2727	Active implant devices/ancillaries	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	25/01/2021	25/01/2021
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	25/01/2021	25/01/2021
							F12	Serious Injury/ Illness/ Impairment	A070504	Premature Discharge of Battery	25/01/2021	25/01/2021
							F19	Surgical Intervention	A070504	Premature Discharge of Battery	25/01/2021	25/01/2021
					CT2750	Implantable cardiac electrical stimulators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	25/01/2021	25/01/2021
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	25/01/2021	25/01/2021
							F12	Serious Injury/ Illness/ Impairment	A070504	Premature Discharge of Battery	25/01/2021	25/01/2021
							F19	Surgical Intervention	A070504	Premature Discharge of Battery	25/01/2021	25/01/2021
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	25/01/2021	25/01/2021
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	25/01/2021	25/01/2021
							F12	Serious Injury/ Illness/ Impairment	A070504	Premature Discharge of Battery	25/01/2021	25/01/2021
							F19	Surgical Intervention	A070504	Premature Discharge of Battery	25/01/2021	25/01/2021
					CT2985	Active	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	25/01/2021	25/01/2021

							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	25/01/2021	25/01/2021
							F12	Serious Injury/ Illness/ Impairment	A070504	Premature Discharge of Battery	25/01/2021	25/01/2021
							F19	Surgical Intervention	A070504	Premature Discharge of Battery	25/01/2021	25/01/2021
					CT3001	Implantable metal	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	25/01/2021	25/01/2021
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	25/01/2021	25/01/2021
							F12	Serious Injury/ Illness/ Impairment	A070504	Premature Discharge of Battery	25/01/2021	25/01/2021
							F19	Surgical Intervention	A070504	Premature Discharge of Battery	25/01/2021	25/01/2021
					CT3002	Implantable synthetic polymer	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	25/01/2021	25/01/2021
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	25/01/2021	25/01/2021
							F12	Serious Injury/ Illness/ Impairment	A070504	Premature Discharge of Battery	25/01/2021	25/01/2021
							F19	Surgical Intervention	A070504	Premature Discharge of Battery	25/01/2021	25/01/2021
					CT3006	Implantable inorganic material	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	25/01/2021	25/01/2021
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	25/01/2021	25/01/2021
							F12	Serious Injury/ Illness/ Impairment	A070504	Premature Discharge of Battery	25/01/2021	25/01/2021
							F19	Surgical Intervention	A070504	Premature Discharge of Battery	25/01/2021	25/01/2021
					CT314	Electrically powered	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	25/01/2021	25/01/2021
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	25/01/2021	25/01/2021
							F12	Serious Injury/ Illness/ Impairment	A070504	Premature Discharge of Battery	25/01/2021	25/01/2021
							F19	Surgical Intervention	A070504	Premature Discharge of Battery	25/01/2021	25/01/2021
					CT321	Long-term surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	25/01/2021	25/01/2021
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	25/01/2021	25/01/2021
							F12	Serious Injury/ Illness/ Impairment	A070504	Premature Discharge of Battery	25/01/2021	25/01/2021
							F19	Surgical Intervention	A070504	Premature Discharge of Battery	25/01/2021	25/01/2021
					CT449	Generators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	25/01/2021	25/01/2021
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	25/01/2021	25/01/2021
							F12	Serious Injury/ Illness/ Impairment	A070504	Premature Discharge of Battery	25/01/2021	25/01/2021
							F19	Surgical Intervention	A070504	Premature Discharge of Battery	25/01/2021	25/01/2021
					CT981	Single use	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	25/01/2021	25/01/2021
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	25/01/2021	25/01/2021
							F12	Serious Injury/ Illness/ Impairment	A070504	Premature Discharge of Battery	25/01/2021	25/01/2021
							F19	Surgical Intervention	A070504	Premature Discharge of Battery	25/01/2021	25/01/2021
					CT983	Surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	25/01/2021	25/01/2021
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	25/01/2021	25/01/2021
							F12	Serious Injury/ Illness/ Impairment	A070504	Premature Discharge of Battery	25/01/2021	25/01/2021
							F19	Surgical Intervention	A070504	Premature Discharge of Battery	25/01/2021	25/01/2021
10533	2021	All other reportable incidents	47265	Dual-chamber implantable pacemaker, rate-responsive	CT141	Electrodes/Leads and associated devices	E2312	Fatigue	A0705	Battery Problem	15/01/2021	15/01/2021
									A070504	Premature Discharge of Battery	15/01/2021	15/01/2021
							F08	Hospitalization or Prolonged Hospitalization	A0705	Battery Problem	15/01/2021	15/01/2021
									A070504	Premature Discharge of Battery	15/01/2021	15/01/2021
							F14	Prolonged Episode of Care	A0705	Battery Problem	15/01/2021	15/01/2021
									A070504	Premature Discharge of Battery	15/01/2021	15/01/2021
							F1903	Device Explantation	A0705	Battery Problem	15/01/2021	15/01/2021
									A070504	Premature Discharge of Battery	15/01/2021	15/01/2021
							F1905	Device Revision or Replacement	A0705	Battery Problem	15/01/2021	15/01/2021
									A070504	Premature Discharge of Battery	15/01/2021	15/01/2021
					CT145	Cardiovascular devices	E2312	Fatigue	A0705	Battery Problem	15/01/2021	15/01/2021
									A070504	Premature Discharge of Battery	15/01/2021	15/01/2021
							F08	Hospitalization or Prolonged Hospitalization	A0705	Battery Problem	15/01/2021	15/01/2021
									A070504	Premature Discharge of Battery	15/01/2021	15/01/2021
							F14	Prolonged Episode of Care	A0705	Battery Problem	15/01/2021	15/01/2021
									A070504	Premature Discharge of Battery	15/01/2021	15/01/2021
							F1903	Device Explantation	A0705	Battery Problem	15/01/2021	15/01/2021
									A070504	Premature Discharge of Battery	15/01/2021	15/01/2021
							F1905	Device Revision or Replacement	A0705	Battery Problem	15/01/2021	15/01/2021
									A070504	Premature Discharge of Battery	15/01/2021	15/01/2021
					CT1684	Cardiac pacemakers and associated devices	E2312	Fatigue	A0705	Battery Problem	15/01/2021	15/01/2021
									A070504	Premature Discharge of Battery	15/01/2021	15/01/2021
							F08	Hospitalization or Prolonged Hospitalization	A0705	Battery Problem	15/01/2021	15/01/2021
									A070504	Premature Discharge of Battery	15/01/2021	15/01/2021
							F14	Prolonged Episode of Care	A0705	Battery Problem	15/01/2021	15/01/2021
									A070504	Premature Discharge of Battery	15/01/2021	15/01/2021
							F1903	Device Explantation	A0705	Battery Problem	15/01/2021	15/01/2021
									A070504	Premature Discharge of Battery	15/01/2021	15/01/2021
							F1905	Device Revision or Replacement	A0705	Battery Problem	15/01/2021	15/01/2021
									A070504	Premature Discharge of Battery	15/01/2021	15/01/2021
					CT1685	Cardiac pacemakers	E2312	Fatigue	A0705	Battery Problem	15/01/2021	15/01/2021
									A070504	Premature Discharge of Battery	15/01/2021	15/01/2021
							F08	Hospitalization or Prolonged Hospitalization	A0705	Battery Problem	15/01/2021	15/01/2021
									A070504	Premature Discharge of Battery	15/01/2021	15/01/2021
							F14	Prolonged Episode of Care	A0705	Battery Problem	15/01/2021	15/01/2021
									A070504	Premature Discharge of Battery	15/01/2021	15/01/2021
							F1903	Device Explantation	A0705	Battery Problem	15/01/2021	15/01/2021
									A070504	Premature Discharge of Battery	15/01/2021	15/01/2021
							F1905	Device Revision or Replacement	A0705	Battery Problem	15/01/2021	15/01/2021
									A070504	Premature Discharge of Battery	15/01/2021	15/01/2021
					CT233	Surgical	E2312	Fatigue	A0705	Battery Problem	15/01/2021	15/01/2021
									A070504	Premature Discharge of Battery	15/01/2021	15/01/2021
							F08	Hospitalization or Prolonged Hospitalization	A0705	Battery Problem	15/01/2021	15/01/2021
									A070504	Premature Discharge of Battery	15/01/2021	15/01/2021
							F14	Prolonged Episode of Care	A0705	Battery Problem	15/01/2021	15/01/2021
									A070504	Premature Discharge of Battery	15/01/2021	15/01/2021
							F1903	Device Explantation	A0705	Battery Problem	15/01/2021	15/01/2021
									A070504	Premature Discharge of Battery	15/01/2021	15/01/2021
							F1905	Device Revision or Replacement	A0705	Battery Problem	15/01/2021	15/01/2021
									A070504	Premature Discharge of Battery	15/01/2021	15/01/2021
					CT2406	Implantable	E2312	Fatigue	A0705	Battery Problem	15/01/2021	15/01/2021
									A070504	Premature Discharge of Battery	15/01/2021	15/01/2021
							F08	Hospitalization or Prolonged Hospitalization	A0705	Battery Problem	15/01/2021	15/01/2021

								A070504	Premature Discharge of Battery	15/01/2021	15/01/2021
						F14	Prolonged Episode of Care	A0705	Battery Problem	15/01/2021	15/01/2021
								A070504	Premature Discharge of Battery	15/01/2021	15/01/2021
						F1903	Device Explantation	A0705	Battery Problem	15/01/2021	15/01/2021
								A070504	Premature Discharge of Battery	15/01/2021	15/01/2021
						F1905	Device Revision or Replacement	A0705	Battery Problem	15/01/2021	15/01/2021
								A070504	Premature Discharge of Battery	15/01/2021	15/01/2021
					CT2407	Active implantable	E2312	Fatigue	A0705	Battery Problem	15/01/2021	15/01/2021
									A070504	Premature Discharge of Battery	15/01/2021	15/01/2021
							F08	Hospitalization or Prolonged Hospitalization	A0705	Battery Problem	15/01/2021	15/01/2021
									A070504	Premature Discharge of Battery	15/01/2021	15/01/2021
							F14	Prolonged Episode of Care	A0705	Battery Problem	15/01/2021	15/01/2021
									A070504	Premature Discharge of Battery	15/01/2021	15/01/2021
							F1903	Device Explantation	A0705	Battery Problem	15/01/2021	15/01/2021
									A070504	Premature Discharge of Battery	15/01/2021	15/01/2021
							F1905	Device Revision or Replacement	A0705	Battery Problem	15/01/2021	15/01/2021
									A070504	Premature Discharge of Battery	15/01/2021	15/01/2021
					CT257	Implantable cardiac pacemakers	E2312	Fatigue	A0705	Battery Problem	15/01/2021	15/01/2021
									A070504	Premature Discharge of Battery	15/01/2021	15/01/2021
							F08	Hospitalization or Prolonged Hospitalization	A0705	Battery Problem	15/01/2021	15/01/2021
									A070504	Premature Discharge of Battery	15/01/2021	15/01/2021
							F14	Prolonged Episode of Care	A0705	Battery Problem	15/01/2021	15/01/2021
									A070504	Premature Discharge of Battery	15/01/2021	15/01/2021
							F1903	Device Explantation	A0705	Battery Problem	15/01/2021	15/01/2021
									A070504	Premature Discharge of Battery	15/01/2021	15/01/2021
							F1905	Device Revision or Replacement	A0705	Battery Problem	15/01/2021	15/01/2021
									A070504	Premature Discharge of Battery	15/01/2021	15/01/2021
					CT2727	Active implant devices/ancillaries	E2312	Fatigue	A0705	Battery Problem	15/01/2021	15/01/2021
									A070504	Premature Discharge of Battery	15/01/2021	15/01/2021
							F08	Hospitalization or Prolonged Hospitalization	A0705	Battery Problem	15/01/2021	15/01/2021
									A070504	Premature Discharge of Battery	15/01/2021	15/01/2021
							F14	Prolonged Episode of Care	A0705	Battery Problem	15/01/2021	15/01/2021
									A070504	Premature Discharge of Battery	15/01/2021	15/01/2021
							F1903	Device Explantation	A0705	Battery Problem	15/01/2021	15/01/2021
									A070504	Premature Discharge of Battery	15/01/2021	15/01/2021
							F1905	Device Revision or Replacement	A0705	Battery Problem	15/01/2021	15/01/2021
									A070504	Premature Discharge of Battery	15/01/2021	15/01/2021
					CT2750	Implantable cardiac electrical stimulators and associated devices	E2312	Fatigue	A0705	Battery Problem	15/01/2021	15/01/2021
									A070504	Premature Discharge of Battery	15/01/2021	15/01/2021
							F08	Hospitalization or Prolonged Hospitalization	A0705	Battery Problem	15/01/2021	15/01/2021
									A070504	Premature Discharge of Battery	15/01/2021	15/01/2021
							F14	Prolonged Episode of Care	A0705	Battery Problem	15/01/2021	15/01/2021
									A070504	Premature Discharge of Battery	15/01/2021	15/01/2021
							F1903	Device Explantation	A0705	Battery Problem	15/01/2021	15/01/2021
									A070504	Premature Discharge of Battery	15/01/2021	15/01/2021
							F1905	Device Revision or Replacement	A0705	Battery Problem	15/01/2021	15/01/2021
									A070504	Premature Discharge of Battery	15/01/2021	15/01/2021
					CT2866	Implantable cardiac electrical stimulators	E2312	Fatigue	A0705	Battery Problem	15/01/2021	15/01/2021
									A070504	Premature Discharge of Battery	15/01/2021	15/01/2021
							F08	Hospitalization or Prolonged Hospitalization	A0705	Battery Problem	15/01/2021	15/01/2021
									A070504	Premature Discharge of Battery	15/01/2021	15/01/2021
							F14	Prolonged Episode of Care	A0705	Battery Problem	15/01/2021	15/01/2021
									A070504	Premature Discharge of Battery	15/01/2021	15/01/2021
							F1903	Device Explantation	A0705	Battery Problem	15/01/2021	15/01/2021
									A070504	Premature Discharge of Battery	15/01/2021	15/01/2021
							F1905	Device Revision or Replacement	A0705	Battery Problem	15/01/2021	15/01/2021
									A070504	Premature Discharge of Battery	15/01/2021	15/01/2021
					CT2985	Active	E2312	Fatigue	A0705	Battery Problem	15/01/2021	15/01/2021
									A070504	Premature Discharge of Battery	15/01/2021	15/01/2021
							F08	Hospitalization or Prolonged Hospitalization	A0705	Battery Problem	15/01/2021	15/01/2021
									A070504	Premature Discharge of Battery	15/01/2021	15/01/2021
							F14	Prolonged Episode of Care	A0705	Battery Problem	15/01/2021	15/01/2021
									A070504	Premature Discharge of Battery	15/01/2021	15/01/2021
							F1903	Device Explantation	A0705	Battery Problem	15/01/2021	15/01/2021
									A070504	Premature Discharge of Battery	15/01/2021	15/01/2021
							F1905	Device Revision or Replacement	A0705	Battery Problem	15/01/2021	15/01/2021
									A070504	Premature Discharge of Battery	15/01/2021	15/01/2021
					CT3001	Implantable metal	E2312	Fatigue	A0705	Battery Problem	15/01/2021	15/01/2021
									A070504	Premature Discharge of Battery	15/01/2021	15/01/2021
							F08	Hospitalization or Prolonged Hospitalization	A0705	Battery Problem	15/01/2021	15/01/2021
									A070504	Premature Discharge of Battery	15/01/2021	15/01/2021
							F14	Prolonged Episode of Care	A0705	Battery Problem	15/01/2021	15/01/2021
									A070504	Premature Discharge of Battery	15/01/2021	15/01/2021
							F1903	Device Explantation	A0705	Battery Problem	15/01/2021	15/01/2021
									A070504	Premature Discharge of Battery	15/01/2021	15/01/2021
							F1905	Device Revision or Replacement	A0705	Battery Problem	15/01/2021	15/01/2021
									A070504	Premature Discharge of Battery	15/01/2021	15/01/2021
					CT3002	Implantable synthetic polymer	E2312	Fatigue	A0705	Battery Problem	15/01/2021	15/01/2021
									A070504	Premature Discharge of Battery	15/01/2021	15/01/2021
							F08	Hospitalization or Prolonged Hospitalization	A0705	Battery Problem	15/01/2021	15/01/2021
									A070504	Premature Discharge of Battery	15/01/2021	15/01/2021
							F14	Prolonged Episode of Care	A0705	Battery Problem	15/01/2021	15/01/2021
									A070504	Premature Discharge of Battery	15/01/2021	15/01/2021
							F1903	Device Explantation	A0705	Battery Problem	15/01/2021	15/01/2021
									A070504	Premature Discharge of Battery	15/01/2021	15/01/2021
							F1905	Device Revision or Replacement	A0705	Battery Problem	15/01/2021	15/01/2021
									A070504	Premature Discharge of Battery	15/01/2021	15/01/2021
					CT3006	Implantable inorganic material	E2312	Fatigue	A0705	Battery Problem	15/01/2021	15/01/2021

									A070504	Premature Discharge of Battery	15/01/2021	15/01/2021
							F08	Hospitalization or Prolonged Hospitalization	A0705	Battery Problem	15/01/2021	15/01/2021
									A070504	Premature Discharge of Battery	15/01/2021	15/01/2021
							F14	Prolonged Episode of Care	A0705	Battery Problem	15/01/2021	15/01/2021
									A070504	Premature Discharge of Battery	15/01/2021	15/01/2021
							F1903	Device Explantation	A0705	Battery Problem	15/01/2021	15/01/2021
									A070504	Premature Discharge of Battery	15/01/2021	15/01/2021
							F1905	Device Revision or Replacement	A0705	Battery Problem	15/01/2021	15/01/2021
									A070504	Premature Discharge of Battery	15/01/2021	15/01/2021
					CT314	Electrically powered	E2312	Fatigue	A0705	Battery Problem	15/01/2021	15/01/2021
									A070504	Premature Discharge of Battery	15/01/2021	15/01/2021
							F08	Hospitalization or Prolonged Hospitalization	A0705	Battery Problem	15/01/2021	15/01/2021
									A070504	Premature Discharge of Battery	15/01/2021	15/01/2021
							F14	Prolonged Episode of Care	A0705	Battery Problem	15/01/2021	15/01/2021
									A070504	Premature Discharge of Battery	15/01/2021	15/01/2021
							F1903	Device Explantation	A0705	Battery Problem	15/01/2021	15/01/2021
									A070504	Premature Discharge of Battery	15/01/2021	15/01/2021
							F1905	Device Revision or Replacement	A0705	Battery Problem	15/01/2021	15/01/2021
									A070504	Premature Discharge of Battery	15/01/2021	15/01/2021
					CT321	Long-term surgical invasive	E2312	Fatigue	A0705	Battery Problem	15/01/2021	15/01/2021
									A070504	Premature Discharge of Battery	15/01/2021	15/01/2021
							F08	Hospitalization or Prolonged Hospitalization	A0705	Battery Problem	15/01/2021	15/01/2021
									A070504	Premature Discharge of Battery	15/01/2021	15/01/2021
							F14	Prolonged Episode of Care	A0705	Battery Problem	15/01/2021	15/01/2021
									A070504	Premature Discharge of Battery	15/01/2021	15/01/2021
							F1903	Device Explantation	A0705	Battery Problem	15/01/2021	15/01/2021
									A070504	Premature Discharge of Battery	15/01/2021	15/01/2021
							F1905	Device Revision or Replacement	A0705	Battery Problem	15/01/2021	15/01/2021
									A070504	Premature Discharge of Battery	15/01/2021	15/01/2021
					CT449	Generators and associated devices	E2312	Fatigue	A0705	Battery Problem	15/01/2021	15/01/2021
									A070504	Premature Discharge of Battery	15/01/2021	15/01/2021
							F08	Hospitalization or Prolonged Hospitalization	A0705	Battery Problem	15/01/2021	15/01/2021
									A070504	Premature Discharge of Battery	15/01/2021	15/01/2021
							F14	Prolonged Episode of Care	A0705	Battery Problem	15/01/2021	15/01/2021
									A070504	Premature Discharge of Battery	15/01/2021	15/01/2021
							F1903	Device Explantation	A0705	Battery Problem	15/01/2021	15/01/2021
									A070504	Premature Discharge of Battery	15/01/2021	15/01/2021
							F1905	Device Revision or Replacement	A0705	Battery Problem	15/01/2021	15/01/2021
									A070504	Premature Discharge of Battery	15/01/2021	15/01/2021
					CT981	Single use	E2312	Fatigue	A0705	Battery Problem	15/01/2021	15/01/2021
									A070504	Premature Discharge of Battery	15/01/2021	15/01/2021
							F08	Hospitalization or Prolonged Hospitalization	A0705	Battery Problem	15/01/2021	15/01/2021
									A070504	Premature Discharge of Battery	15/01/2021	15/01/2021
							F14	Prolonged Episode of Care	A0705	Battery Problem	15/01/2021	15/01/2021
									A070504	Premature Discharge of Battery	15/01/2021	15/01/2021
							F1903	Device Explantation	A0705	Battery Problem	15/01/2021	15/01/2021
									A070504	Premature Discharge of Battery	15/01/2021	15/01/2021
							F1905	Device Revision or Replacement	A0705	Battery Problem	15/01/2021	15/01/2021
									A070504	Premature Discharge of Battery	15/01/2021	15/01/2021
					CT983	Surgical invasive	E2312	Fatigue	A0705	Battery Problem	15/01/2021	15/01/2021
									A070504	Premature Discharge of Battery	15/01/2021	15/01/2021
							F08	Hospitalization or Prolonged Hospitalization	A0705	Battery Problem	15/01/2021	15/01/2021
									A070504	Premature Discharge of Battery	15/01/2021	15/01/2021
							F14	Prolonged Episode of Care	A0705	Battery Problem	15/01/2021	15/01/2021
									A070504	Premature Discharge of Battery	15/01/2021	15/01/2021
							F1903	Device Explantation	A0705	Battery Problem	15/01/2021	15/01/2021
									A070504	Premature Discharge of Battery	15/01/2021	15/01/2021
							F1905	Device Revision or Replacement	A0705	Battery Problem	15/01/2021	15/01/2021
									A070504	Premature Discharge of Battery	15/01/2021	15/01/2021
10534	2021	All other reportable incidents	47265	Dual-chamber implantable pacemaker, rate-responsive	CT141	Electrodes/Leads and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	19/03/2021	19/03/2021
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	19/03/2021	19/03/2021
							F1903	Device Explantation	A070504	Premature Discharge of Battery	19/03/2021	19/03/2021
					CT145	Cardiovascular devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	19/03/2021	19/03/2021
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	19/03/2021	19/03/2021
							F1903	Device Explantation	A070504	Premature Discharge of Battery	19/03/2021	19/03/2021
					CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	19/03/2021	19/03/2021
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	19/03/2021	19/03/2021
							F1903	Device Explantation	A070504	Premature Discharge of Battery	19/03/2021	19/03/2021
					CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	19/03/2021	19/03/2021
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	19/03/2021	19/03/2021
							F1903	Device Explantation	A070504	Premature Discharge of Battery	19/03/2021	19/03/2021
					CT233	Surgical	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	19/03/2021	19/03/2021
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	19/03/2021	19/03/2021
							F1903	Device Explantation	A070504	Premature Discharge of Battery	19/03/2021	19/03/2021
					CT2406	Implantable	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	19/03/2021	19/03/2021
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	19/03/2021	19/03/2021
							F1903	Device Explantation	A070504	Premature Discharge of Battery	19/03/2021	19/03/2021
					CT2407	Active implantable	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	19/03/2021	19/03/2021
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	19/03/2021	19/03/2021
							F1903	Device Explantation	A070504	Premature Discharge of Battery	19/03/2021	19/03/2021
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	19/03/2021	19/03/2021
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	19/03/2021	19/03/2021
							F1903	Device Explantation	A070504	Premature Discharge of Battery	19/03/2021	19/03/2021
					CT2727	Active implant devices/ancillaries	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	19/03/2021	19/03/2021
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	19/03/2021	19/03/2021
							F1903	Device Explantation	A070504	Premature Discharge of Battery	19/03/2021	19/03/2021
					CT2750	Implantable cardiac electrical stimulators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	19/03/2021	19/03/2021
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	19/03/2021	19/03/2021

							F1903	Device Explantation	A070504	Premature Discharge of Battery	19/03/2021	19/03/2021
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	19/03/2021	19/03/2021
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	19/03/2021	19/03/2021
							F1903	Device Explantation	A070504	Premature Discharge of Battery	19/03/2021	19/03/2021
					CT2985	Active	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	19/03/2021	19/03/2021
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	19/03/2021	19/03/2021
							F1903	Device Explantation	A070504	Premature Discharge of Battery	19/03/2021	19/03/2021
					CT3001	Implantable metal	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	19/03/2021	19/03/2021
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	19/03/2021	19/03/2021
							F1903	Device Explantation	A070504	Premature Discharge of Battery	19/03/2021	19/03/2021
					CT3002	Implantable synthetic polymer	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	19/03/2021	19/03/2021
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	19/03/2021	19/03/2021
							F1903	Device Explantation	A070504	Premature Discharge of Battery	19/03/2021	19/03/2021
					CT3006	Implantable inorganic material	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	19/03/2021	19/03/2021
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	19/03/2021	19/03/2021
							F1903	Device Explantation	A070504	Premature Discharge of Battery	19/03/2021	19/03/2021
					CT314	Electrically powered	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	19/03/2021	19/03/2021
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	19/03/2021	19/03/2021
							F1903	Device Explantation	A070504	Premature Discharge of Battery	19/03/2021	19/03/2021
					CT321	Long-term surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	19/03/2021	19/03/2021
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	19/03/2021	19/03/2021
							F1903	Device Explantation	A070504	Premature Discharge of Battery	19/03/2021	19/03/2021
					CT449	Generators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	19/03/2021	19/03/2021
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	19/03/2021	19/03/2021
							F1903	Device Explantation	A070504	Premature Discharge of Battery	19/03/2021	19/03/2021
					CT981	Single use	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	19/03/2021	19/03/2021
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	19/03/2021	19/03/2021
							F1903	Device Explantation	A070504	Premature Discharge of Battery	19/03/2021	19/03/2021
					CT983	Surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	19/03/2021	19/03/2021
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	19/03/2021	19/03/2021
							F1903	Device Explantation	A070504	Premature Discharge of Battery	19/03/2021	19/03/2021
10535	2021	All other reportable incidents	47270	Cardiac resynchronization therapy implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A070909	Oversensing		
							F1905	Device Revision or Replacement	A070909	Oversensing		
							F26	No Health Consequences or Impact	A070909	Oversensing		
10536	2021	All other reportable incidents	47266	Single-chamber implantable pacemaker, fixed-rate/demand	CT141	Electrodes/Leads and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	21/01/2021	21/01/2021
									A09	Output Problem	21/01/2021	21/01/2021
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	21/01/2021	21/01/2021
									A09	Output Problem	21/01/2021	21/01/2021
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery	21/01/2021	21/01/2021
									A09	Output Problem	21/01/2021	21/01/2021
					CT145	Cardiovascular devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	21/01/2021	21/01/2021
									A09	Output Problem	21/01/2021	21/01/2021
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	21/01/2021	21/01/2021
									A09	Output Problem	21/01/2021	21/01/2021
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery	21/01/2021	21/01/2021
									A09	Output Problem	21/01/2021	21/01/2021
					CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	21/01/2021	21/01/2021
									A09	Output Problem	21/01/2021	21/01/2021
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	21/01/2021	21/01/2021
									A09	Output Problem	21/01/2021	21/01/2021
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery	21/01/2021	21/01/2021
									A09	Output Problem	21/01/2021	21/01/2021
					CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	21/01/2021	21/01/2021
									A09	Output Problem	21/01/2021	21/01/2021
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	21/01/2021	21/01/2021
									A09	Output Problem	21/01/2021	21/01/2021
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery	21/01/2021	21/01/2021
									A09	Output Problem	21/01/2021	21/01/2021
					CT233	Surgical	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	21/01/2021	21/01/2021
									A09	Output Problem	21/01/2021	21/01/2021
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	21/01/2021	21/01/2021
									A09	Output Problem	21/01/2021	21/01/2021
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery	21/01/2021	21/01/2021
									A09	Output Problem	21/01/2021	21/01/2021
					CT2406	Implantable	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	21/01/2021	21/01/2021
									A09	Output Problem	21/01/2021	21/01/2021
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	21/01/2021	21/01/2021
									A09	Output Problem	21/01/2021	21/01/2021
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery	21/01/2021	21/01/2021
									A09	Output Problem	21/01/2021	21/01/2021
					CT2407	Active implantable	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	21/01/2021	21/01/2021
									A09	Output Problem	21/01/2021	21/01/2021
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	21/01/2021	21/01/2021
									A09	Output Problem	21/01/2021	21/01/2021
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery	21/01/2021	21/01/2021
									A09	Output Problem	21/01/2021	21/01/2021
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	21/01/2021	21/01/2021
									A09	Output Problem	21/01/2021	21/01/2021
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	21/01/2021	21/01/2021
									A09	Output Problem	21/01/2021	21/01/2021
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery	21/01/2021	21/01/2021
									A09	Output Problem	21/01/2021	21/01/2021
					CT2727	Active implant devices/ancillaries	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	21/01/2021	21/01/2021
									A09	Output Problem	21/01/2021	21/01/2021
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	21/01/2021	21/01/2021
									A09	Output Problem	21/01/2021	21/01/2021
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery	21/01/2021	21/01/2021
									A09	Output Problem	21/01/2021	21/01/2021

					CT2750	Implantable cardiac electrical stimulators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	21/01/2021	21/01/2021
									A09	Output Problem	21/01/2021	21/01/2021
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	21/01/2021	21/01/2021
									A09	Output Problem	21/01/2021	21/01/2021
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery	21/01/2021	21/01/2021
									A09	Output Problem	21/01/2021	21/01/2021
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	21/01/2021	21/01/2021
									A09	Output Problem	21/01/2021	21/01/2021
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	21/01/2021	21/01/2021
									A09	Output Problem	21/01/2021	21/01/2021
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery	21/01/2021	21/01/2021
									A09	Output Problem	21/01/2021	21/01/2021
					CT2985	Active	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	21/01/2021	21/01/2021
									A09	Output Problem	21/01/2021	21/01/2021
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	21/01/2021	21/01/2021
									A09	Output Problem	21/01/2021	21/01/2021
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery	21/01/2021	21/01/2021
									A09	Output Problem	21/01/2021	21/01/2021
					CT3001	Implantable metal	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	21/01/2021	21/01/2021
									A09	Output Problem	21/01/2021	21/01/2021
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	21/01/2021	21/01/2021
									A09	Output Problem	21/01/2021	21/01/2021
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery	21/01/2021	21/01/2021
									A09	Output Problem	21/01/2021	21/01/2021
					CT3002	Implantable synthetic polymer	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	21/01/2021	21/01/2021
									A09	Output Problem	21/01/2021	21/01/2021
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	21/01/2021	21/01/2021
									A09	Output Problem	21/01/2021	21/01/2021
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery	21/01/2021	21/01/2021
									A09	Output Problem	21/01/2021	21/01/2021
					CT3006	Implantable inorganic material	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	21/01/2021	21/01/2021
									A09	Output Problem	21/01/2021	21/01/2021
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	21/01/2021	21/01/2021
									A09	Output Problem	21/01/2021	21/01/2021
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery	21/01/2021	21/01/2021
									A09	Output Problem	21/01/2021	21/01/2021
					CT314	Electrically powered	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	21/01/2021	21/01/2021
									A09	Output Problem	21/01/2021	21/01/2021
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	21/01/2021	21/01/2021
									A09	Output Problem	21/01/2021	21/01/2021
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery	21/01/2021	21/01/2021
									A09	Output Problem	21/01/2021	21/01/2021
					CT321	Long-term surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	21/01/2021	21/01/2021
									A09	Output Problem	21/01/2021	21/01/2021
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	21/01/2021	21/01/2021
									A09	Output Problem	21/01/2021	21/01/2021
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery	21/01/2021	21/01/2021
									A09	Output Problem	21/01/2021	21/01/2021
					CT449	Generators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	21/01/2021	21/01/2021
									A09	Output Problem	21/01/2021	21/01/2021
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	21/01/2021	21/01/2021
									A09	Output Problem	21/01/2021	21/01/2021
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery	21/01/2021	21/01/2021
									A09	Output Problem	21/01/2021	21/01/2021
					CT981	Single use	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	21/01/2021	21/01/2021
									A09	Output Problem	21/01/2021	21/01/2021
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	21/01/2021	21/01/2021
									A09	Output Problem	21/01/2021	21/01/2021
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery	21/01/2021	21/01/2021
									A09	Output Problem	21/01/2021	21/01/2021
					CT983	Surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	21/01/2021	21/01/2021
									A09	Output Problem	21/01/2021	21/01/2021
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	21/01/2021	21/01/2021
									A09	Output Problem	21/01/2021	21/01/2021
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery	21/01/2021	21/01/2021
									A09	Output Problem	21/01/2021	21/01/2021
10537	2021	All other reportable incidents	35852	Single-chamber implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem	29/01/2021	29/01/2021
									A070504	Premature Discharge of Battery	29/01/2021	29/01/2021
							F19	Surgical Intervention	A0705	Battery Problem	29/01/2021	29/01/2021
									A070504	Premature Discharge of Battery	29/01/2021	29/01/2021
10538	2021	All other reportable incidents	17882	Physiologic-monitoring defibrillation system	CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070803	Failure to Power Up		
							F27	Problem Identified During Non-Clinical Procedure	A070803	Failure to Power Up		
10539	2021	All other reportable incidents	17882	Physiologic-monitoring defibrillation system	CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070803	Failure to Power Up		
							F27	Problem Identified During Non-Clinical Procedure	A070803	Failure to Power Up		
10540	2021	All other reportable incidents	17882	Physiologic-monitoring defibrillation system	CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070803	Failure to Power Up		
							F1001	Absence of Treatment	A070803	Failure to Power Up		
							F26	No Health Consequences or Impact	A070803	Failure to Power Up		
10541	2021	All other reportable incidents	17882	Physiologic-monitoring defibrillation system	CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070803	Failure to Power Up		
							F27	Problem Identified During Non-Clinical Procedure	A070803	Failure to Power Up		
10542	2021	All other reportable incidents	17882	Physiologic-monitoring defibrillation system	CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070803	Failure to Power Up		
							F27	Problem Identified During Non-Clinical Procedure	A070803	Failure to Power Up		
10543	2021	All other reportable incidents	35852	Single-chamber implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem	01/03/2021	01/03/2021
									A070504	Premature Discharge of Battery	01/03/2021	01/03/2021
							F19	Surgical Intervention	A0705	Battery Problem	01/03/2021	01/03/2021
									A070504	Premature Discharge of Battery	01/03/2021	01/03/2021
10544	2021	All other reportable incidents	17882	Physiologic-monitoring defibrillation system	CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A1601	Device Alarm System		
							F26	No Health Consequences or Impact	A1601	Device Alarm System		
10545	2021	All other reportable incidents	47265	Dual-chamber implantable pacemaker, rate-	CT141	Electrodes/Leads and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	21/01/2021	21/01/2021

				responsive					A090805	Incorrect Measurement	21/01/2021	21/01/2021
							F1901	Additional Surgery	A070504	Premature Discharge of Battery	21/01/2021	21/01/2021
									A090805	Incorrect Measurement	21/01/2021	21/01/2021
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	21/01/2021	21/01/2021
									A090805	Incorrect Measurement	21/01/2021	21/01/2021
					CT145	Cardiovascular devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	21/01/2021	21/01/2021
									A090805	Incorrect Measurement	21/01/2021	21/01/2021
							F1901	Additional Surgery	A070504	Premature Discharge of Battery	21/01/2021	21/01/2021
									A090805	Incorrect Measurement	21/01/2021	21/01/2021
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	21/01/2021	21/01/2021
									A090805	Incorrect Measurement	21/01/2021	21/01/2021
					CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	21/01/2021	21/01/2021
									A090805	Incorrect Measurement	21/01/2021	21/01/2021
							F1901	Additional Surgery	A070504	Premature Discharge of Battery	21/01/2021	21/01/2021
									A090805	Incorrect Measurement	21/01/2021	21/01/2021
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	21/01/2021	21/01/2021
									A090805	Incorrect Measurement	21/01/2021	21/01/2021
					CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	21/01/2021	21/01/2021
									A090805	Incorrect Measurement	21/01/2021	21/01/2021
							F1901	Additional Surgery	A070504	Premature Discharge of Battery	21/01/2021	21/01/2021
									A090805	Incorrect Measurement	21/01/2021	21/01/2021
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	21/01/2021	21/01/2021
									A090805	Incorrect Measurement	21/01/2021	21/01/2021
					CT233	Surgical	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	21/01/2021	21/01/2021
									A090805	Incorrect Measurement	21/01/2021	21/01/2021
							F1901	Additional Surgery	A070504	Premature Discharge of Battery	21/01/2021	21/01/2021
									A090805	Incorrect Measurement	21/01/2021	21/01/2021
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	21/01/2021	21/01/2021
									A090805	Incorrect Measurement	21/01/2021	21/01/2021
					CT2406	Implantable	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	21/01/2021	21/01/2021
									A090805	Incorrect Measurement	21/01/2021	21/01/2021
							F1901	Additional Surgery	A070504	Premature Discharge of Battery	21/01/2021	21/01/2021
									A090805	Incorrect Measurement	21/01/2021	21/01/2021
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	21/01/2021	21/01/2021
									A090805	Incorrect Measurement	21/01/2021	21/01/2021
					CT2407	Active implantable	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	21/01/2021	21/01/2021
									A090805	Incorrect Measurement	21/01/2021	21/01/2021
							F1901	Additional Surgery	A070504	Premature Discharge of Battery	21/01/2021	21/01/2021
									A090805	Incorrect Measurement	21/01/2021	21/01/2021
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	21/01/2021	21/01/2021
									A090805	Incorrect Measurement	21/01/2021	21/01/2021
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	21/01/2021	21/01/2021
									A090805	Incorrect Measurement	21/01/2021	21/01/2021
							F1901	Additional Surgery	A070504	Premature Discharge of Battery	21/01/2021	21/01/2021
									A090805	Incorrect Measurement	21/01/2021	21/01/2021
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	21/01/2021	21/01/2021
									A090805	Incorrect Measurement	21/01/2021	21/01/2021
					CT2727	Active implant devices/ancillaries	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	21/01/2021	21/01/2021
									A090805	Incorrect Measurement	21/01/2021	21/01/2021
							F1901	Additional Surgery	A070504	Premature Discharge of Battery	21/01/2021	21/01/2021
									A090805	Incorrect Measurement	21/01/2021	21/01/2021
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	21/01/2021	21/01/2021
									A090805	Incorrect Measurement	21/01/2021	21/01/2021
					CT2750	Implantable cardiac electrical stimulators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	21/01/2021	21/01/2021
									A090805	Incorrect Measurement	21/01/2021	21/01/2021
							F1901	Additional Surgery	A070504	Premature Discharge of Battery	21/01/2021	21/01/2021
									A090805	Incorrect Measurement	21/01/2021	21/01/2021
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	21/01/2021	21/01/2021
									A090805	Incorrect Measurement	21/01/2021	21/01/2021
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	21/01/2021	21/01/2021
									A090805	Incorrect Measurement	21/01/2021	21/01/2021
							F1901	Additional Surgery	A070504	Premature Discharge of Battery	21/01/2021	21/01/2021
									A090805	Incorrect Measurement	21/01/2021	21/01/2021
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	21/01/2021	21/01/2021
									A090805	Incorrect Measurement	21/01/2021	21/01/2021
					CT2985	Active	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	21/01/2021	21/01/2021
									A090805	Incorrect Measurement	21/01/2021	21/01/2021
							F1901	Additional Surgery	A070504	Premature Discharge of Battery	21/01/2021	21/01/2021
									A090805	Incorrect Measurement	21/01/2021	21/01/2021
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	21/01/2021	21/01/2021
									A090805	Incorrect Measurement	21/01/2021	21/01/2021
					CT3001	Implantable metal	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	21/01/2021	21/01/2021
									A090805	Incorrect Measurement	21/01/2021	21/01/2021
							F1901	Additional Surgery	A070504	Premature Discharge of Battery	21/01/2021	21/01/2021
									A090805	Incorrect Measurement	21/01/2021	21/01/2021
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	21/01/2021	21/01/2021
									A090805	Incorrect Measurement	21/01/2021	21/01/2021
					CT3002	Implantable synthetic polymer	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	21/01/2021	21/01/2021
									A090805	Incorrect Measurement	21/01/2021	21/01/2021
							F1901	Additional Surgery	A070504	Premature Discharge of Battery	21/01/2021	21/01/2021
									A090805	Incorrect Measurement	21/01/2021	21/01/2021
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	21/01/2021	21/01/2021
									A090805	Incorrect Measurement	21/01/2021	21/01/2021
					CT3006	Implantable inorganic material	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	21/01/2021	21/01/2021
									A090805	Incorrect Measurement	21/01/2021	21/01/2021
							F1901	Additional Surgery	A070504	Premature Discharge of Battery	21/01/2021	21/01/2021
									A090805	Incorrect Measurement	21/01/2021	21/01/2021
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	21/01/2021	21/01/2021

									A090805	Incorrect Measurement	21/01/2021	21/01/2021
					CT314	Electrically powered	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	21/01/2021	21/01/2021
									A090805	Incorrect Measurement	21/01/2021	21/01/2021
							F1901	Additional Surgery	A070504	Premature Discharge of Battery	21/01/2021	21/01/2021
									A090805	Incorrect Measurement	21/01/2021	21/01/2021
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	21/01/2021	21/01/2021
									A090805	Incorrect Measurement	21/01/2021	21/01/2021
					CT321	Long-term surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	21/01/2021	21/01/2021
									A090805	Incorrect Measurement	21/01/2021	21/01/2021
							F1901	Additional Surgery	A070504	Premature Discharge of Battery	21/01/2021	21/01/2021
									A090805	Incorrect Measurement	21/01/2021	21/01/2021
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	21/01/2021	21/01/2021
									A090805	Incorrect Measurement	21/01/2021	21/01/2021
					CT449	Generators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	21/01/2021	21/01/2021
									A090805	Incorrect Measurement	21/01/2021	21/01/2021
							F1901	Additional Surgery	A070504	Premature Discharge of Battery	21/01/2021	21/01/2021
									A090805	Incorrect Measurement	21/01/2021	21/01/2021
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	21/01/2021	21/01/2021
									A090805	Incorrect Measurement	21/01/2021	21/01/2021
					CT981	Single use	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	21/01/2021	21/01/2021
									A090805	Incorrect Measurement	21/01/2021	21/01/2021
							F1901	Additional Surgery	A070504	Premature Discharge of Battery	21/01/2021	21/01/2021
									A090805	Incorrect Measurement	21/01/2021	21/01/2021
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	21/01/2021	21/01/2021
									A090805	Incorrect Measurement	21/01/2021	21/01/2021
					CT983	Surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	21/01/2021	21/01/2021
									A090805	Incorrect Measurement	21/01/2021	21/01/2021
							F1901	Additional Surgery	A070504	Premature Discharge of Battery	21/01/2021	21/01/2021
									A090805	Incorrect Measurement	21/01/2021	21/01/2021
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	21/01/2021	21/01/2021
									A090805	Incorrect Measurement	21/01/2021	21/01/2021
10546	2021	All other reportable incidents	35223	Endocardial/interventricular septal pacing lead	CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A0401	Break		
									A070909	Oversensing		
							F26	No Health Consequences or Impact	A0401	Break		
									A070909	Oversensing		
10547	2021	All other reportable incidents	47267	Single-chamber implantable pacemaker, rate-responsive	CT141	Electrodes/Leads and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
					CT145	Cardiovascular devices	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
					CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
					CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
					CT233	Surgical	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
					CT2406	Implantable	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
					CT2407	Active implantable	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
					CT2727	Active implant devices/ancillaries	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
					CT2750	Implantable cardiac electrical stimulators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
					CT2985	Active	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
					CT3001	Implantable metal	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
					CT3002	Implantable synthetic polymer	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem		

									A070504	Premature Discharge of Battery		
						F26	No Health Consequences or Impact		A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
					CT3006	Implantable inorganic material	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
					CT314	Electrically powered	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
					CT321	Long-term surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
					CT449	Generators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
					CT981	Single use	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
					CT983	Surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
10548	2021	All other reportable incidents	35852	Single-chamber implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem	27/01/2021	27/01/2021
									A070504	Premature Discharge of Battery	27/01/2021	27/01/2021
							F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A0705	Battery Problem	27/01/2021	27/01/2021
									A070504	Premature Discharge of Battery	27/01/2021	27/01/2021
							F19	Surgical Intervention	A0705	Battery Problem	27/01/2021	27/01/2021
									A070504	Premature Discharge of Battery	27/01/2021	27/01/2021
10549	2021	All other reportable incidents	47265	Dual-chamber implantable pacemaker, rate-responsive	CT141	Electrodes/Leads and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
							F24	Insufficient Health Impact Information	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
					CT145	Cardiovascular devices	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
							F24	Insufficient Health Impact Information	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
					CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
							F24	Insufficient Health Impact Information	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
					CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
							F24	Insufficient Health Impact Information	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
					CT233	Surgical	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
							F24	Insufficient Health Impact Information	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
					CT2406	Implantable	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
							F24	Insufficient Health Impact Information	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
					CT2407	Active implantable	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
							F24	Insufficient Health Impact Information	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
							F24	Insufficient Health Impact Information	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
					CT2727	Active implant devices/ancillaries	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
							F24	Insufficient Health Impact Information	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
					CT2750	Implantable cardiac electrical stimulators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
							F24	Insufficient Health Impact Information	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
							F24	Insufficient Health Impact Information	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
					CT2985	Active	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
							F24	Insufficient Health Impact Information	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
					CT3001	Implantable metal	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
							F24	Insufficient Health Impact Information	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
					CT3002	Implantable synthetic polymer	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
							F24	Insufficient Health Impact Information	A0705	Battery Problem		

									A070504	Premature Discharge of Battery		
				CT3006	Implantable inorganic material	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem			
								A070504	Premature Discharge of Battery			
						F24	Insufficient Health Impact Information	A0705	Battery Problem			
								A070504	Premature Discharge of Battery			
				CT314	Electrically powered	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem			
								A070504	Premature Discharge of Battery			
						F24	Insufficient Health Impact Information	A0705	Battery Problem			
								A070504	Premature Discharge of Battery			
				CT321	Long-term surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem			
								A070504	Premature Discharge of Battery			
						F24	Insufficient Health Impact Information	A0705	Battery Problem			
								A070504	Premature Discharge of Battery			
				CT449	Generators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem			
								A070504	Premature Discharge of Battery			
						F24	Insufficient Health Impact Information	A0705	Battery Problem			
								A070504	Premature Discharge of Battery			
				CT981	Single use	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem			
								A070504	Premature Discharge of Battery			
						F24	Insufficient Health Impact Information	A0705	Battery Problem			
								A070504	Premature Discharge of Battery			
				CT983	Surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem			
								A070504	Premature Discharge of Battery			
						F24	Insufficient Health Impact Information	A0705	Battery Problem			
								A070504	Premature Discharge of Battery			
10550	2021	All other reportable incidents	36241	Pacemaker/defibrillator lead	CT1684	Cardiac pacemakers and associated devices	E2114	Perforation	A051201	Device Dislodged or Dislocated	27/11/2020	27/11/2020
							F08	Hospitalization or Prolonged Hospitalization	A051201	Device Dislodged or Dislocated	27/11/2020	27/11/2020
							F12	Serious Injury/ Illness/ Impairment	A051201	Device Dislodged or Dislocated	27/11/2020	27/11/2020
							F19	Surgical Intervention	A051201	Device Dislodged or Dislocated	27/11/2020	27/11/2020
							F1905	Device Revision or Replacement	A051201	Device Dislodged or Dislocated	27/11/2020	27/11/2020
10551	2021	All other reportable incidents	17882	Physiologic-monitoring defibrillation system	CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A1603	Failure of Device to Self-Test		
							F27	Problem Identified During Non-Clinical Procedure	A1603	Failure of Device to Self-Test		
10552	2021	All other reportable incidents	47267	Single-chamber implantable pacemaker, rate-responsive	CT141	Electrodes/Leads and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	15/01/2021	15/01/2021
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	15/01/2021	15/01/2021
					CT145	Cardiovascular devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	15/01/2021	15/01/2021
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	15/01/2021	15/01/2021
					CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	15/01/2021	15/01/2021
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	15/01/2021	15/01/2021
					CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	15/01/2021	15/01/2021
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	15/01/2021	15/01/2021
					CT233	Surgical	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	15/01/2021	15/01/2021
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	15/01/2021	15/01/2021
					CT2406	Implantable	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	15/01/2021	15/01/2021
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	15/01/2021	15/01/2021
					CT2407	Active implantable	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	15/01/2021	15/01/2021
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	15/01/2021	15/01/2021
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	15/01/2021	15/01/2021
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	15/01/2021	15/01/2021
					CT2727	Active implant devices/ancillaries	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	15/01/2021	15/01/2021
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	15/01/2021	15/01/2021
					CT2750	Implantable cardiac electrical stimulators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	15/01/2021	15/01/2021
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	15/01/2021	15/01/2021
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	15/01/2021	15/01/2021
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	15/01/2021	15/01/2021
					CT2985	Active	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	15/01/2021	15/01/2021
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	15/01/2021	15/01/2021
					CT3001	Implantable metal	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	15/01/2021	15/01/2021
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	15/01/2021	15/01/2021
					CT3002	Implantable synthetic polymer	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	15/01/2021	15/01/2021
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	15/01/2021	15/01/2021
					CT3006	Implantable inorganic material	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	15/01/2021	15/01/2021
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	15/01/2021	15/01/2021
					CT314	Electrically powered	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	15/01/2021	15/01/2021
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	15/01/2021	15/01/2021
					CT321	Long-term surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	15/01/2021	15/01/2021
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	15/01/2021	15/01/2021
					CT449	Generators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	15/01/2021	15/01/2021
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	15/01/2021	15/01/2021
					CT981	Single use	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	15/01/2021	15/01/2021
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	15/01/2021	15/01/2021
					CT983	Surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	15/01/2021	15/01/2021
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	15/01/2021	15/01/2021
10553	2021	All other reportable incidents	35223	Endocardial/interventricular septal pacing lead	CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A0701	Capturing Problem		
									A072202	Low Impedance		
							F26	No Health Consequences or Impact	A0701	Capturing Problem		
									A072202	Low Impedance		
10554	2021	All other reportable incidents	35852	Single-chamber implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	09/03/2021	09/03/2021
							F19	Surgical Intervention	A070504	Premature Discharge of Battery	09/03/2021	09/03/2021
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	09/03/2021	09/03/2021
10555	2021	All other reportable incidents	36241	Pacemaker/defibrillator lead	CT1684	Cardiac pacemakers and associated devices	E2104	Electric Shock	A070909	Oversensing		
									A071302	Inappropriate/Inadequate Shock/Stimulation		
							E2403	No Clinical Signs, Symptoms or Conditions	A070909	Oversensing		
									A071302	Inappropriate/Inadequate Shock/Stimulation		
							F26	No Health Consequences or Impact	A070909	Oversensing		
									A071302	Inappropriate/Inadequate Shock/Stimulation		
10556	2021	All other reportable incidents	35852	Single-chamber implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E2401	Insufficient Information	A070909	Oversensing		
							E2403	No Clinical Signs, Symptoms or Conditions	A070909	Oversensing		
							F23	Unexpected Medical Intervention	A070909	Oversensing		

10557	2021	All other reportable incidents	36241	Pacemaker/defibrillator lead	CT1684	Cardiac pacemakers and associated devices	E060104	Bradycardia	A070909	Oversensing		
									A0712	Pacing Problem		
									A090801	Signal Artifact/Noise		
							F08	Hospitalization or Prolonged Hospitalization	A070909	Oversensing		
									A0712	Pacing Problem		
									A090801	Signal Artifact/Noise		
							F12	Serious Injury/ Illness/ Impairment	A070909	Oversensing		
									A0712	Pacing Problem		
									A090801	Signal Artifact/Noise		
							F1905	Device Revision or Replacement	A070909	Oversensing		
									A0712	Pacing Problem		
									A090801	Signal Artifact/Noise		
10558	2021	All other reportable incidents	37265	Dual-chamber implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
10559	2021	All other reportable incidents	47267	Single-chamber implantable pacemaker, rate-responsive	CT141	Electrodes/Leads and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	11/01/2021	11/01/2021
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	11/01/2021	11/01/2021
							F12	Serious Injury/ Illness/ Impairment	A070504	Premature Discharge of Battery	11/01/2021	11/01/2021
							F19	Surgical Intervention	A070504	Premature Discharge of Battery	11/01/2021	11/01/2021
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	11/01/2021	11/01/2021
					CT145	Cardiovascular devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	11/01/2021	11/01/2021
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	11/01/2021	11/01/2021
							F12	Serious Injury/ Illness/ Impairment	A070504	Premature Discharge of Battery	11/01/2021	11/01/2021
							F19	Surgical Intervention	A070504	Premature Discharge of Battery	11/01/2021	11/01/2021
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	11/01/2021	11/01/2021
					CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	11/01/2021	11/01/2021
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	11/01/2021	11/01/2021
							F12	Serious Injury/ Illness/ Impairment	A070504	Premature Discharge of Battery	11/01/2021	11/01/2021
							F19	Surgical Intervention	A070504	Premature Discharge of Battery	11/01/2021	11/01/2021
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	11/01/2021	11/01/2021
					CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	11/01/2021	11/01/2021
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	11/01/2021	11/01/2021
							F12	Serious Injury/ Illness/ Impairment	A070504	Premature Discharge of Battery	11/01/2021	11/01/2021
							F19	Surgical Intervention	A070504	Premature Discharge of Battery	11/01/2021	11/01/2021
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	11/01/2021	11/01/2021
					CT233	Surgical	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	11/01/2021	11/01/2021
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	11/01/2021	11/01/2021
							F12	Serious Injury/ Illness/ Impairment	A070504	Premature Discharge of Battery	11/01/2021	11/01/2021
							F19	Surgical Intervention	A070504	Premature Discharge of Battery	11/01/2021	11/01/2021
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	11/01/2021	11/01/2021
					CT2406	Implantable	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	11/01/2021	11/01/2021
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	11/01/2021	11/01/2021
							F12	Serious Injury/ Illness/ Impairment	A070504	Premature Discharge of Battery	11/01/2021	11/01/2021
							F19	Surgical Intervention	A070504	Premature Discharge of Battery	11/01/2021	11/01/2021
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	11/01/2021	11/01/2021
					CT2407	Active implantable	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	11/01/2021	11/01/2021
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	11/01/2021	11/01/2021
							F12	Serious Injury/ Illness/ Impairment	A070504	Premature Discharge of Battery	11/01/2021	11/01/2021
							F19	Surgical Intervention	A070504	Premature Discharge of Battery	11/01/2021	11/01/2021
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	11/01/2021	11/01/2021
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	11/01/2021	11/01/2021
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	11/01/2021	11/01/2021
							F12	Serious Injury/ Illness/ Impairment	A070504	Premature Discharge of Battery	11/01/2021	11/01/2021
							F19	Surgical Intervention	A070504	Premature Discharge of Battery	11/01/2021	11/01/2021
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	11/01/2021	11/01/2021
					CT2727	Active implant devices/ancillaries	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	11/01/2021	11/01/2021
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	11/01/2021	11/01/2021
							F12	Serious Injury/ Illness/ Impairment	A070504	Premature Discharge of Battery	11/01/2021	11/01/2021
							F19	Surgical Intervention	A070504	Premature Discharge of Battery	11/01/2021	11/01/2021
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	11/01/2021	11/01/2021
					CT2750	Implantable cardiac electrical stimulators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	11/01/2021	11/01/2021
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	11/01/2021	11/01/2021
							F12	Serious Injury/ Illness/ Impairment	A070504	Premature Discharge of Battery	11/01/2021	11/01/2021
							F19	Surgical Intervention	A070504	Premature Discharge of Battery	11/01/2021	11/01/2021
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	11/01/2021	11/01/2021
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	11/01/2021	11/01/2021
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	11/01/2021	11/01/2021
							F12	Serious Injury/ Illness/ Impairment	A070504	Premature Discharge of Battery	11/01/2021	11/01/2021
							F19	Surgical Intervention	A070504	Premature Discharge of Battery	11/01/2021	11/01/2021
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	11/01/2021	11/01/2021
					CT2985	Active	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	11/01/2021	11/01/2021
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	11/01/2021	11/01/2021
							F12	Serious Injury/ Illness/ Impairment	A070504	Premature Discharge of Battery	11/01/2021	11/01/2021
							F19	Surgical Intervention	A070504	Premature Discharge of Battery	11/01/2021	11/01/2021
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	11/01/2021	11/01/2021
					CT3001	Implantable metal	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	11/01/2021	11/01/2021
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	11/01/2021	11/01/2021
							F12	Serious Injury/ Illness/ Impairment	A070504	Premature Discharge of Battery	11/01/2021	11/01/2021
							F19	Surgical Intervention	A070504	Premature Discharge of Battery	11/01/2021	11/01/2021
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	11/01/2021	11/01/2021
					CT3002	Implantable synthetic polymer	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	11/01/2021	11/01/2021
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	11/01/2021	11/01/2021
							F12	Serious Injury/ Illness/ Impairment	A070504	Premature Discharge of Battery	11/01/2021	11/01/2021
							F19	Surgical Intervention	A070504	Premature Discharge of Battery	11/01/2021	11/01/2021
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	11/01/2021	11/01/2021
					CT3006	Implantable inorganic material	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	11/01/2021	11/01/2021
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	11/01/2021	11/01/2021
							F12	Serious Injury/ Illness/ Impairment	A070504	Premature Discharge of Battery	11/01/2021	11/01/2021
							F19	Surgical Intervention	A070504	Premature Discharge of Battery	11/01/2021	11/01/2021

							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	11/01/2021	11/01/2021
					CT314	Electrically powered	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	11/01/2021	11/01/2021
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	11/01/2021	11/01/2021
							F12	Serious Injury/ Illness/ Impairment	A070504	Premature Discharge of Battery	11/01/2021	11/01/2021
							F19	Surgical Intervention	A070504	Premature Discharge of Battery	11/01/2021	11/01/2021
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	11/01/2021	11/01/2021
					CT321	Long-term surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	11/01/2021	11/01/2021
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	11/01/2021	11/01/2021
							F12	Serious Injury/ Illness/ Impairment	A070504	Premature Discharge of Battery	11/01/2021	11/01/2021
							F19	Surgical Intervention	A070504	Premature Discharge of Battery	11/01/2021	11/01/2021
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	11/01/2021	11/01/2021
					CT449	Generators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	11/01/2021	11/01/2021
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	11/01/2021	11/01/2021
							F12	Serious Injury/ Illness/ Impairment	A070504	Premature Discharge of Battery	11/01/2021	11/01/2021
							F19	Surgical Intervention	A070504	Premature Discharge of Battery	11/01/2021	11/01/2021
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	11/01/2021	11/01/2021
					CT981	Single use	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	11/01/2021	11/01/2021
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	11/01/2021	11/01/2021
							F12	Serious Injury/ Illness/ Impairment	A070504	Premature Discharge of Battery	11/01/2021	11/01/2021
							F19	Surgical Intervention	A070504	Premature Discharge of Battery	11/01/2021	11/01/2021
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	11/01/2021	11/01/2021
					CT983	Surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	11/01/2021	11/01/2021
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	11/01/2021	11/01/2021
							F12	Serious Injury/ Illness/ Impairment	A070504	Premature Discharge of Battery	11/01/2021	11/01/2021
							F19	Surgical Intervention	A070504	Premature Discharge of Battery	11/01/2021	11/01/2021
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	11/01/2021	11/01/2021
10560	2021	All other reportable incidents	35852	Single-chamber implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	15/03/2021	15/03/2021
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	15/03/2021	15/03/2021
							F12	Serious Injury/ Illness/ Impairment	A070504	Premature Discharge of Battery	15/03/2021	15/03/2021
							F19	Surgical Intervention	A070504	Premature Discharge of Battery	15/03/2021	15/03/2021
10561	2021	All other reportable incidents	47267	Single-chamber implantable pacemaker, rate-responsive	CT141	Electrodes/Leads and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT145	Cardiovascular devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT233	Surgical	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT2406	Implantable	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT2407	Active implantable	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT2727	Active implant devices/ancillaries	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT2750	Implantable cardiac electrical stimulators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT2985	Active	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT3001	Implantable metal	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT3002	Implantable synthetic polymer	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT3006	Implantable inorganic material	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT314	Electrically powered	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT321	Long-term surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT449	Generators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT981	Single use	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT983	Surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
10562	2021	All other reportable incidents	35852	Single-chamber implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F24	Insufficient Health Impact Information	A070504	Premature Discharge of Battery		
10563	2021	All other reportable incidents	35852	Single-chamber implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E2104	Electric Shock	A070909	Oversensing		
									A071302	Inappropriate/Inadequate Shock/Stimulation		
							E2403	No Clinical Signs, Symptoms or Conditions	A070909	Oversensing		
									A071302	Inappropriate/Inadequate Shock/Stimulation		
							F26	No Health Consequences or Impact	A070909	Oversensing		
									A071302	Inappropriate/Inadequate Shock/Stimulation		
10564	2021	All other reportable incidents	47265	Dual-chamber implantable pacemaker, rate-responsive	CT141	Electrodes/Leads and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT145	Cardiovascular devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT233	Surgical	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		

					CT2406	Implantable	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT2407	Active implantable	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT2727	Active implant devices/ancillaries	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT2750	Implantable cardiac electrical stimulators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT2985	Active	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT3001	Implantable metal	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT3002	Implantable synthetic polymer	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT3006	Implantable inorganic material	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT314	Electrically powered	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT321	Long-term surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT449	Generators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT981	Single use	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT983	Surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
10565	2021	All other reportable incidents	47267	Single-chamber implantable pacemaker, rate-responsive	CT141	Electrodes/Leads and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A072201	High Impedance		
							F26	No Health Consequences or Impact	A072201	High Impedance		
					CT145	Cardiovascular devices	E2403	No Clinical Signs, Symptoms or Conditions	A072201	High Impedance		
							F26	No Health Consequences or Impact	A072201	High Impedance		
					CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A072201	High Impedance		
							F26	No Health Consequences or Impact	A072201	High Impedance		
					CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A072201	High Impedance		
							F26	No Health Consequences or Impact	A072201	High Impedance		
					CT233	Surgical	E2403	No Clinical Signs, Symptoms or Conditions	A072201	High Impedance		
							F26	No Health Consequences or Impact	A072201	High Impedance		
					CT2406	Implantable	E2403	No Clinical Signs, Symptoms or Conditions	A072201	High Impedance		
							F26	No Health Consequences or Impact	A072201	High Impedance		
					CT2407	Active implantable	E2403	No Clinical Signs, Symptoms or Conditions	A072201	High Impedance		
							F26	No Health Consequences or Impact	A072201	High Impedance		
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A072201	High Impedance		
							F26	No Health Consequences or Impact	A072201	High Impedance		
					CT2727	Active implant devices/ancillaries	E2403	No Clinical Signs, Symptoms or Conditions	A072201	High Impedance		
							F26	No Health Consequences or Impact	A072201	High Impedance		
					CT2750	Implantable cardiac electrical stimulators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A072201	High Impedance		
							F26	No Health Consequences or Impact	A072201	High Impedance		
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A072201	High Impedance		
							F26	No Health Consequences or Impact	A072201	High Impedance		
					CT2985	Active	E2403	No Clinical Signs, Symptoms or Conditions	A072201	High Impedance		
							F26	No Health Consequences or Impact	A072201	High Impedance		
					CT3001	Implantable metal	E2403	No Clinical Signs, Symptoms or Conditions	A072201	High Impedance		
							F26	No Health Consequences or Impact	A072201	High Impedance		
					CT3002	Implantable synthetic polymer	E2403	No Clinical Signs, Symptoms or Conditions	A072201	High Impedance		
							F26	No Health Consequences or Impact	A072201	High Impedance		
					CT3006	Implantable inorganic material	E2403	No Clinical Signs, Symptoms or Conditions	A072201	High Impedance		
							F26	No Health Consequences or Impact	A072201	High Impedance		
					CT314	Electrically powered	E2403	No Clinical Signs, Symptoms or Conditions	A072201	High Impedance		
							F26	No Health Consequences or Impact	A072201	High Impedance		
					CT321	Long-term surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A072201	High Impedance		
							F26	No Health Consequences or Impact	A072201	High Impedance		
					CT449	Generators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A072201	High Impedance		
							F26	No Health Consequences or Impact	A072201	High Impedance		
					CT981	Single use	E2403	No Clinical Signs, Symptoms or Conditions	A072201	High Impedance		
							F26	No Health Consequences or Impact	A072201	High Impedance		
					CT983	Surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A072201	High Impedance		
							F26	No Health Consequences or Impact	A072201	High Impedance		
10566	2021	All other reportable incidents	35039	Epicardial pacing lead	CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A040101	Fracture		
									A090801	Signal Artifact/Noise		
							F26	No Health Consequences or Impact	A040101	Fracture		
									A090801	Signal Artifact/Noise		
10567	2021	All other reportable incidents	35852	Single-chamber implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem	15/02/2021	15/02/2021
									A070504	Premature Discharge of Battery	15/02/2021	15/02/2021
							F19	Surgical Intervention	A0705	Battery Problem	15/02/2021	15/02/2021
									A070504	Premature Discharge of Battery	15/02/2021	15/02/2021
10568	2021	All other reportable incidents	47265	Dual-chamber implantable pacemaker, rate-responsive	CT141	Electrodes/Leads and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	26/05/2021	26/05/2021
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	26/05/2021	26/05/2021
							F19	Surgical Intervention	A070504	Premature Discharge of Battery	26/05/2021	26/05/2021
					CT145	Cardiovascular devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	26/05/2021	26/05/2021
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	26/05/2021	26/05/2021
							F19	Surgical Intervention	A070504	Premature Discharge of Battery	26/05/2021	26/05/2021
					CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	26/05/2021	26/05/2021
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	26/05/2021	26/05/2021
							F19	Surgical Intervention	A070504	Premature Discharge of Battery	26/05/2021	26/05/2021
					CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	26/05/2021	26/05/2021

							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	26/05/2021	26/05/2021
							F19	Surgical Intervention	A070504	Premature Discharge of Battery	26/05/2021	26/05/2021
					CT233	Surgical	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	26/05/2021	26/05/2021
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	26/05/2021	26/05/2021
							F19	Surgical Intervention	A070504	Premature Discharge of Battery	26/05/2021	26/05/2021
					CT2406	Implantable	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	26/05/2021	26/05/2021
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	26/05/2021	26/05/2021
							F19	Surgical Intervention	A070504	Premature Discharge of Battery	26/05/2021	26/05/2021
					CT2407	Active implantable	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	26/05/2021	26/05/2021
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	26/05/2021	26/05/2021
							F19	Surgical Intervention	A070504	Premature Discharge of Battery	26/05/2021	26/05/2021
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	26/05/2021	26/05/2021
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	26/05/2021	26/05/2021
							F19	Surgical Intervention	A070504	Premature Discharge of Battery	26/05/2021	26/05/2021
					CT2727	Active implant devices/ancillaries	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	26/05/2021	26/05/2021
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	26/05/2021	26/05/2021
							F19	Surgical Intervention	A070504	Premature Discharge of Battery	26/05/2021	26/05/2021
					CT2750	Implantable cardiac electrical stimulators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	26/05/2021	26/05/2021
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	26/05/2021	26/05/2021
							F19	Surgical Intervention	A070504	Premature Discharge of Battery	26/05/2021	26/05/2021
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	26/05/2021	26/05/2021
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	26/05/2021	26/05/2021
							F19	Surgical Intervention	A070504	Premature Discharge of Battery	26/05/2021	26/05/2021
					CT2985	Active	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	26/05/2021	26/05/2021
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	26/05/2021	26/05/2021
							F19	Surgical Intervention	A070504	Premature Discharge of Battery	26/05/2021	26/05/2021
					CT3001	Implantable metal	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	26/05/2021	26/05/2021
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	26/05/2021	26/05/2021
							F19	Surgical Intervention	A070504	Premature Discharge of Battery	26/05/2021	26/05/2021
					CT3002	Implantable synthetic polymer	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	26/05/2021	26/05/2021
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	26/05/2021	26/05/2021
							F19	Surgical Intervention	A070504	Premature Discharge of Battery	26/05/2021	26/05/2021
					CT3006	Implantable inorganic material	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	26/05/2021	26/05/2021
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	26/05/2021	26/05/2021
							F19	Surgical Intervention	A070504	Premature Discharge of Battery	26/05/2021	26/05/2021
					CT314	Electrically powered	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	26/05/2021	26/05/2021
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	26/05/2021	26/05/2021
							F19	Surgical Intervention	A070504	Premature Discharge of Battery	26/05/2021	26/05/2021
					CT321	Long-term surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	26/05/2021	26/05/2021
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	26/05/2021	26/05/2021
							F19	Surgical Intervention	A070504	Premature Discharge of Battery	26/05/2021	26/05/2021
					CT449	Generators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	26/05/2021	26/05/2021
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	26/05/2021	26/05/2021
							F19	Surgical Intervention	A070504	Premature Discharge of Battery	26/05/2021	26/05/2021
					CT981	Single use	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	26/05/2021	26/05/2021
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	26/05/2021	26/05/2021
							F19	Surgical Intervention	A070504	Premature Discharge of Battery	26/05/2021	26/05/2021
					CT983	Surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	26/05/2021	26/05/2021
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	26/05/2021	26/05/2021
							F19	Surgical Intervention	A070504	Premature Discharge of Battery	26/05/2021	26/05/2021
10569	2021	All other reportable incidents	35852	Single-chamber implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A070909	Oversensing		
									A070910	Undersensing		
							F26	No Health Consequences or Impact	A070909	Oversensing		
									A070910	Undersensing		
10570	2021	All other reportable incidents	35852	Single-chamber implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A070909	Oversensing		
							F26	No Health Consequences or Impact	A070909	Oversensing		
10571	2021	All other reportable incidents	37265	Dual-chamber implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	15/03/2021	15/03/2021
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	15/03/2021	15/03/2021
							F12	Serious Injury/ Illness/ Impairment	A070504	Premature Discharge of Battery	15/03/2021	15/03/2021
							F19	Surgical Intervention	A070504	Premature Discharge of Battery	15/03/2021	15/03/2021
10572	2021	All other reportable incidents	47270	Cardiac resynchronization therapy implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E2104	Electric Shock	A071302	Inappropriate/Inadequate Shock/Stimulation		
							E2403	No Clinical Signs, Symptoms or Conditions	A071302	Inappropriate/Inadequate Shock/Stimulation		
							F26	No Health Consequences or Impact	A071302	Inappropriate/Inadequate Shock/Stimulation		
10573	2021	All other reportable incidents	47267	Single-chamber implantable pacemaker, rate-responsive	CT141	Electrodes/Leads and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	06/03/2021	07/03/2021
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	06/03/2021	07/03/2021
							F12	Serious Injury/ Illness/ Impairment	A070504	Premature Discharge of Battery	06/03/2021	07/03/2021
							F19	Surgical Intervention	A070504	Premature Discharge of Battery	06/03/2021	07/03/2021
					CT145	Cardiovascular devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	06/03/2021	07/03/2021
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	06/03/2021	07/03/2021
							F12	Serious Injury/ Illness/ Impairment	A070504	Premature Discharge of Battery	06/03/2021	07/03/2021
							F19	Surgical Intervention	A070504	Premature Discharge of Battery	06/03/2021	07/03/2021
					CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	06/03/2021	07/03/2021
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	06/03/2021	07/03/2021
							F12	Serious Injury/ Illness/ Impairment	A070504	Premature Discharge of Battery	06/03/2021	07/03/2021
							F19	Surgical Intervention	A070504	Premature Discharge of Battery	06/03/2021	07/03/2021
					CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	06/03/2021	07/03/2021
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	06/03/2021	07/03/2021
							F12	Serious Injury/ Illness/ Impairment	A070504	Premature Discharge of Battery	06/03/2021	07/03/2021
							F19	Surgical Intervention	A070504	Premature Discharge of Battery	06/03/2021	07/03/2021
					CT233	Surgical	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	06/03/2021	07/03/2021
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	06/03/2021	07/03/2021
							F12	Serious Injury/ Illness/ Impairment	A070504	Premature Discharge of Battery	06/03/2021	07/03/2021
							F19	Surgical Intervention	A070504	Premature Discharge of Battery	06/03/2021	07/03/2021
					CT2406	Implantable	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	06/03/2021	07/03/2021
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	06/03/2021	07/03/2021
							F12	Serious Injury/ Illness/ Impairment	A070504	Premature Discharge of Battery	06/03/2021	07/03/2021
							F19	Surgical Intervention	A070504	Premature Discharge of Battery	06/03/2021	07/03/2021
					CT2407	Active implantable	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	06/03/2021	07/03/2021

							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	06/03/2021	07/03/2021
							F12	Serious Injury/ Illness/ Impairment	A070504	Premature Discharge of Battery	06/03/2021	07/03/2021
							F19	Surgical Intervention	A070504	Premature Discharge of Battery	06/03/2021	07/03/2021
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	06/03/2021	07/03/2021
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	06/03/2021	07/03/2021
							F12	Serious Injury/ Illness/ Impairment	A070504	Premature Discharge of Battery	06/03/2021	07/03/2021
							F19	Surgical Intervention	A070504	Premature Discharge of Battery	06/03/2021	07/03/2021
					CT2727	Active implant devices/ancillaries	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	06/03/2021	07/03/2021
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	06/03/2021	07/03/2021
							F12	Serious Injury/ Illness/ Impairment	A070504	Premature Discharge of Battery	06/03/2021	07/03/2021
							F19	Surgical Intervention	A070504	Premature Discharge of Battery	06/03/2021	07/03/2021
					CT2750	Implantable cardiac electrical stimulators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	06/03/2021	07/03/2021
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	06/03/2021	07/03/2021
							F12	Serious Injury/ Illness/ Impairment	A070504	Premature Discharge of Battery	06/03/2021	07/03/2021
							F19	Surgical Intervention	A070504	Premature Discharge of Battery	06/03/2021	07/03/2021
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	06/03/2021	07/03/2021
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	06/03/2021	07/03/2021
							F12	Serious Injury/ Illness/ Impairment	A070504	Premature Discharge of Battery	06/03/2021	07/03/2021
							F19	Surgical Intervention	A070504	Premature Discharge of Battery	06/03/2021	07/03/2021
					CT2985	Active	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	06/03/2021	07/03/2021
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	06/03/2021	07/03/2021
							F12	Serious Injury/ Illness/ Impairment	A070504	Premature Discharge of Battery	06/03/2021	07/03/2021
							F19	Surgical Intervention	A070504	Premature Discharge of Battery	06/03/2021	07/03/2021
					CT3001	Implantable metal	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	06/03/2021	07/03/2021
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	06/03/2021	07/03/2021
							F12	Serious Injury/ Illness/ Impairment	A070504	Premature Discharge of Battery	06/03/2021	07/03/2021
							F19	Surgical Intervention	A070504	Premature Discharge of Battery	06/03/2021	07/03/2021
					CT3002	Implantable synthetic polymer	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	06/03/2021	07/03/2021
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	06/03/2021	07/03/2021
							F12	Serious Injury/ Illness/ Impairment	A070504	Premature Discharge of Battery	06/03/2021	07/03/2021
							F19	Surgical Intervention	A070504	Premature Discharge of Battery	06/03/2021	07/03/2021
					CT3006	Implantable inorganic material	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	06/03/2021	07/03/2021
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	06/03/2021	07/03/2021
							F12	Serious Injury/ Illness/ Impairment	A070504	Premature Discharge of Battery	06/03/2021	07/03/2021
							F19	Surgical Intervention	A070504	Premature Discharge of Battery	06/03/2021	07/03/2021
					CT314	Electrically powered	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	06/03/2021	07/03/2021
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	06/03/2021	07/03/2021
							F12	Serious Injury/ Illness/ Impairment	A070504	Premature Discharge of Battery	06/03/2021	07/03/2021
							F19	Surgical Intervention	A070504	Premature Discharge of Battery	06/03/2021	07/03/2021
					CT321	Long-term surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	06/03/2021	07/03/2021
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	06/03/2021	07/03/2021
							F12	Serious Injury/ Illness/ Impairment	A070504	Premature Discharge of Battery	06/03/2021	07/03/2021
							F19	Surgical Intervention	A070504	Premature Discharge of Battery	06/03/2021	07/03/2021
					CT449	Generators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	06/03/2021	07/03/2021
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	06/03/2021	07/03/2021
							F12	Serious Injury/ Illness/ Impairment	A070504	Premature Discharge of Battery	06/03/2021	07/03/2021
							F19	Surgical Intervention	A070504	Premature Discharge of Battery	06/03/2021	07/03/2021
					CT981	Single use	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	06/03/2021	07/03/2021
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	06/03/2021	07/03/2021
							F12	Serious Injury/ Illness/ Impairment	A070504	Premature Discharge of Battery	06/03/2021	07/03/2021
							F19	Surgical Intervention	A070504	Premature Discharge of Battery	06/03/2021	07/03/2021
					CT983	Surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	06/03/2021	07/03/2021
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	06/03/2021	07/03/2021
							F12	Serious Injury/ Illness/ Impairment	A070504	Premature Discharge of Battery	06/03/2021	07/03/2021
							F19	Surgical Intervention	A070504	Premature Discharge of Battery	06/03/2021	07/03/2021
10574	2021	All other reportable incidents	17882	Physiologic-monitoring defibrillation system	CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A07	Electrical /Electronic Property Problem		
							F26	No Health Consequences or Impact	A07	Electrical /Electronic Property Problem		
10575	2021	All other reportable incidents	17882	Physiologic-monitoring defibrillation system	CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A0719	Unexpected Shutdown		
							F27	Problem Identified During Non-Clinical Procedure	A0719	Unexpected Shutdown		
10576	2021	All other reportable incidents	35852	Single-chamber implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem	29/01/2021	29/01/2021
									A070504	Premature Discharge of Battery	29/01/2021	29/01/2021
							F19	Surgical Intervention	A0705	Battery Problem	29/01/2021	29/01/2021
									A070504	Premature Discharge of Battery	29/01/2021	29/01/2021
10577	2021	All other reportable incidents	35852	Single-chamber implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A070909	Oversensing		
							F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A070909	Oversensing		
							F26	No Health Consequences or Impact	A070909	Oversensing		
10578	2021	All other reportable incidents	35852	Single-chamber implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E2104	Electric Shock	A071302	Inappropriate/Inadequate Shock/Stimulation		
							F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A071302	Inappropriate/Inadequate Shock/Stimulation		
							F26	No Health Consequences or Impact	A071302	Inappropriate/Inadequate Shock/Stimulation		
10579	2021	All other reportable incidents	47270	Cardiac resynchronization therapy implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	26/01/2021	26/01/2021
							F1901	Additional Surgery	A070504	Premature Discharge of Battery	26/01/2021	26/01/2021
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	26/01/2021	26/01/2021
10580	2021	All other reportable incidents	47205	Cardiac pulse generator programmer	CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A11	Computer Software Problem		
							F26	No Health Consequences or Impact	A11	Computer Software Problem		
10581	2021	All other reportable incidents	47265	Dual-chamber implantable pacemaker, rate-responsive	CT141	Electrodes/Leads and associated devices	E011903	Syncope/Fainting	A070101	Failure to Capture	09/02/2021	09/02/2021
									A070909	Oversensing	09/02/2021	09/02/2021
									A0712	Pacing Problem	09/02/2021	09/02/2021
									A072201	High Impedance	09/02/2021	09/02/2021
									A090801	Signal Artifact/Noise	09/02/2021	09/02/2021
							E060104	Bradycardia	A070101	Failure to Capture	09/02/2021	09/02/2021
									A070909	Oversensing	09/02/2021	09/02/2021
									A0712	Pacing Problem	09/02/2021	09/02/2021
									A072201	High Impedance	09/02/2021	09/02/2021
									A090801	Signal Artifact/Noise	09/02/2021	09/02/2021
							F08	Hospitalization or Prolonged Hospitalization	A070101	Failure to Capture	09/02/2021	09/02/2021
									A070909	Oversensing	09/02/2021	09/02/2021
									A0712	Pacing Problem	09/02/2021	09/02/2021
									A072201	High Impedance	09/02/2021	09/02/2021

								A090801	Signal Artifact/Noise	09/02/2021	09/02/2021
						F12	Serious Injury/ Illness/ Impairment	A070101	Failure to Capture	09/02/2021	09/02/2021
								A070909	Oversensing	09/02/2021	09/02/2021
								A0712	Pacing Problem	09/02/2021	09/02/2021
								A072201	High Impedance	09/02/2021	09/02/2021
								A090801	Signal Artifact/Noise	09/02/2021	09/02/2021
						F15	Recognized Device or Procedural Complication	A070101	Failure to Capture	09/02/2021	09/02/2021
								A070909	Oversensing	09/02/2021	09/02/2021
								A0712	Pacing Problem	09/02/2021	09/02/2021
								A072201	High Impedance	09/02/2021	09/02/2021
								A090801	Signal Artifact/Noise	09/02/2021	09/02/2021
						F1903	Device Explantation	A070101	Failure to Capture	09/02/2021	09/02/2021
								A070909	Oversensing	09/02/2021	09/02/2021
								A0712	Pacing Problem	09/02/2021	09/02/2021
								A072201	High Impedance	09/02/2021	09/02/2021
								A090801	Signal Artifact/Noise	09/02/2021	09/02/2021
					CT145	Cardiovascular devices	E011903	Syncope/Fainting	A070101	Failure to Capture	09/02/2021	09/02/2021
									A070909	Oversensing	09/02/2021	09/02/2021
									A0712	Pacing Problem	09/02/2021	09/02/2021
									A072201	High Impedance	09/02/2021	09/02/2021
									A090801	Signal Artifact/Noise	09/02/2021	09/02/2021
							E060104	Bradycardia	A070101	Failure to Capture	09/02/2021	09/02/2021
									A070909	Oversensing	09/02/2021	09/02/2021
									A0712	Pacing Problem	09/02/2021	09/02/2021
									A072201	High Impedance	09/02/2021	09/02/2021
									A090801	Signal Artifact/Noise	09/02/2021	09/02/2021
							F08	Hospitalization or Prolonged Hospitalization	A070101	Failure to Capture	09/02/2021	09/02/2021
									A070909	Oversensing	09/02/2021	09/02/2021
									A0712	Pacing Problem	09/02/2021	09/02/2021
									A072201	High Impedance	09/02/2021	09/02/2021
									A090801	Signal Artifact/Noise	09/02/2021	09/02/2021
							F12	Serious Injury/ Illness/ Impairment	A070101	Failure to Capture	09/02/2021	09/02/2021
									A070909	Oversensing	09/02/2021	09/02/2021
									A0712	Pacing Problem	09/02/2021	09/02/2021
									A072201	High Impedance	09/02/2021	09/02/2021
									A090801	Signal Artifact/Noise	09/02/2021	09/02/2021
							F15	Recognized Device or Procedural Complication	A070101	Failure to Capture	09/02/2021	09/02/2021
									A070909	Oversensing	09/02/2021	09/02/2021
									A0712	Pacing Problem	09/02/2021	09/02/2021
									A072201	High Impedance	09/02/2021	09/02/2021
									A090801	Signal Artifact/Noise	09/02/2021	09/02/2021
							F1903	Device Explantation	A070101	Failure to Capture	09/02/2021	09/02/2021
									A070909	Oversensing	09/02/2021	09/02/2021
									A0712	Pacing Problem	09/02/2021	09/02/2021
									A072201	High Impedance	09/02/2021	09/02/2021
									A090801	Signal Artifact/Noise	09/02/2021	09/02/2021
					CT1684	Cardiac pacemakers and associated devices	E011903	Syncope/Fainting	A070101	Failure to Capture	09/02/2021	09/02/2021
									A070909	Oversensing	09/02/2021	09/02/2021
									A0712	Pacing Problem	09/02/2021	09/02/2021
									A072201	High Impedance	09/02/2021	09/02/2021
									A090801	Signal Artifact/Noise	09/02/2021	09/02/2021
							E060104	Bradycardia	A070101	Failure to Capture	09/02/2021	09/02/2021
									A070909	Oversensing	09/02/2021	09/02/2021
									A0712	Pacing Problem	09/02/2021	09/02/2021
									A072201	High Impedance	09/02/2021	09/02/2021
									A090801	Signal Artifact/Noise	09/02/2021	09/02/2021
							F08	Hospitalization or Prolonged Hospitalization	A070101	Failure to Capture	09/02/2021	09/02/2021
									A070909	Oversensing	09/02/2021	09/02/2021
									A0712	Pacing Problem	09/02/2021	09/02/2021
									A072201	High Impedance	09/02/2021	09/02/2021
									A090801	Signal Artifact/Noise	09/02/2021	09/02/2021
							F12	Serious Injury/ Illness/ Impairment	A070101	Failure to Capture	09/02/2021	09/02/2021
									A070909	Oversensing	09/02/2021	09/02/2021
									A0712	Pacing Problem	09/02/2021	09/02/2021
									A072201	High Impedance	09/02/2021	09/02/2021
									A090801	Signal Artifact/Noise	09/02/2021	09/02/2021
							F15	Recognized Device or Procedural Complication	A070101	Failure to Capture	09/02/2021	09/02/2021
									A070909	Oversensing	09/02/2021	09/02/2021
									A0712	Pacing Problem	09/02/2021	09/02/2021
									A072201	High Impedance	09/02/2021	09/02/2021
									A090801	Signal Artifact/Noise	09/02/2021	09/02/2021
							F1903	Device Explantation	A070101	Failure to Capture	09/02/2021	09/02/2021
									A070909	Oversensing	09/02/2021	09/02/2021
									A0712	Pacing Problem	09/02/2021	09/02/2021
									A072201	High Impedance	09/02/2021	09/02/2021
									A090801	Signal Artifact/Noise	09/02/2021	09/02/2021
					CT1685	Cardiac pacemakers	E011903	Syncope/Fainting	A070101	Failure to Capture	09/02/2021	09/02/2021
									A070909	Oversensing	09/02/2021	09/02/2021
									A0712	Pacing Problem	09/02/2021	09/02/2021
									A072201	High Impedance	09/02/2021	09/02/2021
									A090801	Signal Artifact/Noise	09/02/2021	09/02/2021
							E060104	Bradycardia	A070101	Failure to Capture	09/02/2021	09/02/2021
									A070909	Oversensing	09/02/2021	09/02/2021
									A0712	Pacing Problem	09/02/2021	09/02/2021
									A072201	High Impedance	09/02/2021	09/02/2021
									A090801	Signal Artifact/Noise	09/02/2021	09/02/2021
							F08	Hospitalization or Prolonged Hospitalization	A070101	Failure to Capture	09/02/2021	09/02/2021
									A070909	Oversensing	09/02/2021	09/02/2021

								A0712	Pacing Problem	09/02/2021	09/02/2021
								A072201	High Impedance	09/02/2021	09/02/2021
								A090801	Signal Artifact/Noise	09/02/2021	09/02/2021
							F12	Serious Injury/ Illness/ Impairment	A070101	Failure to Capture	09/02/2021	09/02/2021
									A070909	Oversensing	09/02/2021	09/02/2021
									A0712	Pacing Problem	09/02/2021	09/02/2021
									A072201	High Impedance	09/02/2021	09/02/2021
									A090801	Signal Artifact/Noise	09/02/2021	09/02/2021
							F15	Recognized Device or Procedural Complication	A070101	Failure to Capture	09/02/2021	09/02/2021
									A070909	Oversensing	09/02/2021	09/02/2021
									A0712	Pacing Problem	09/02/2021	09/02/2021
									A072201	High Impedance	09/02/2021	09/02/2021
									A090801	Signal Artifact/Noise	09/02/2021	09/02/2021
							F1903	Device Explantation	A070101	Failure to Capture	09/02/2021	09/02/2021
									A070909	Oversensing	09/02/2021	09/02/2021
									A0712	Pacing Problem	09/02/2021	09/02/2021
									A072201	High Impedance	09/02/2021	09/02/2021
									A090801	Signal Artifact/Noise	09/02/2021	09/02/2021
						CT233	Surgical	E011903	Syncope/Fainting	A070101	Failure to Capture	09/02/2021	09/02/2021
										A070909	Oversensing	09/02/2021	09/02/2021
										A0712	Pacing Problem	09/02/2021	09/02/2021
										A072201	High Impedance	09/02/2021	09/02/2021
										A090801	Signal Artifact/Noise	09/02/2021	09/02/2021
								E060104	Bradycardia	A070101	Failure to Capture	09/02/2021	09/02/2021
										A070909	Oversensing	09/02/2021	09/02/2021
										A0712	Pacing Problem	09/02/2021	09/02/2021
										A072201	High Impedance	09/02/2021	09/02/2021
										A090801	Signal Artifact/Noise	09/02/2021	09/02/2021
								F08	Hospitalization or Prolonged Hospitalization	A070101	Failure to Capture	09/02/2021	09/02/2021
										A070909	Oversensing	09/02/2021	09/02/2021
										A0712	Pacing Problem	09/02/2021	09/02/2021
										A072201	High Impedance	09/02/2021	09/02/2021
										A090801	Signal Artifact/Noise	09/02/2021	09/02/2021
								F12	Serious Injury/ Illness/ Impairment	A070101	Failure to Capture	09/02/2021	09/02/2021
										A070909	Oversensing	09/02/2021	09/02/2021
										A0712	Pacing Problem	09/02/2021	09/02/2021
										A072201	High Impedance	09/02/2021	09/02/2021
										A090801	Signal Artifact/Noise	09/02/2021	09/02/2021
								F15	Recognized Device or Procedural Complication	A070101	Failure to Capture	09/02/2021	09/02/2021
										A070909	Oversensing	09/02/2021	09/02/2021
										A0712	Pacing Problem	09/02/2021	09/02/2021
										A072201	High Impedance	09/02/2021	09/02/2021
										A090801	Signal Artifact/Noise	09/02/2021	09/02/2021
								F1903	Device Explantation	A070101	Failure to Capture	09/02/2021	09/02/2021
										A070909	Oversensing	09/02/2021	09/02/2021
										A0712	Pacing Problem	09/02/2021	09/02/2021
										A072201	High Impedance	09/02/2021	09/02/2021
										A090801	Signal Artifact/Noise	09/02/2021	09/02/2021
						CT2406	Implantable	E011903	Syncope/Fainting	A070101	Failure to Capture	09/02/2021	09/02/2021
										A070909	Oversensing	09/02/2021	09/02/2021
										A0712	Pacing Problem	09/02/2021	09/02/2021
										A072201	High Impedance	09/02/2021	09/02/2021
										A090801	Signal Artifact/Noise	09/02/2021	09/02/2021
								E060104	Bradycardia	A070101	Failure to Capture	09/02/2021	09/02/2021
										A070909	Oversensing	09/02/2021	09/02/2021
										A0712	Pacing Problem	09/02/2021	09/02/2021
										A072201	High Impedance	09/02/2021	09/02/2021
										A090801	Signal Artifact/Noise	09/02/2021	09/02/2021
								F08	Hospitalization or Prolonged Hospitalization	A070101	Failure to Capture	09/02/2021	09/02/2021
										A070909	Oversensing	09/02/2021	09/02/2021
										A0712	Pacing Problem	09/02/2021	09/02/2021
										A072201	High Impedance	09/02/2021	09/02/2021
										A090801	Signal Artifact/Noise	09/02/2021	09/02/2021
								F12	Serious Injury/ Illness/ Impairment	A070101	Failure to Capture	09/02/2021	09/02/2021
										A070909	Oversensing	09/02/2021	09/02/2021
										A0712	Pacing Problem	09/02/2021	09/02/2021
										A072201	High Impedance	09/02/2021	09/02/2021
										A090801	Signal Artifact/Noise	09/02/2021	09/02/2021
								F15	Recognized Device or Procedural Complication	A070101	Failure to Capture	09/02/2021	09/02/2021
										A070909	Oversensing	09/02/2021	09/02/2021
										A0712	Pacing Problem	09/02/2021	09/02/2021
										A072201	High Impedance	09/02/2021	09/02/2021
										A090801	Signal Artifact/Noise	09/02/2021	09/02/2021
								F1903	Device Explantation	A070101	Failure to Capture	09/02/2021	09/02/2021
										A070909	Oversensing	09/02/2021	09/02/2021
										A0712	Pacing Problem	09/02/2021	09/02/2021
										A072201	High Impedance	09/02/2021	09/02/2021
										A090801	Signal Artifact/Noise	09/02/2021	09/02/2021
						CT2407	Active implantable	E011903	Syncope/Fainting	A070101	Failure to Capture	09/02/2021	09/02/2021
										A070909	Oversensing	09/02/2021	09/02/2021
										A0712	Pacing Problem	09/02/2021	09/02/2021
										A072201	High Impedance	09/02/2021	09/02/2021
										A090801	Signal Artifact/Noise	09/02/2021	09/02/2021
								E060104	Bradycardia	A070101	Failure to Capture	09/02/2021	09/02/2021
										A070909	Oversensing	09/02/2021	09/02/2021
										A0712	Pacing Problem	09/02/2021	09/02/2021
										A072201	High Impedance	09/02/2021	09/02/2021
										A090801	Signal Artifact/Noise	09/02/2021	09/02/2021

							F08	Hospitalization or Prolonged Hospitalization	A070101	Failure to Capture	09/02/2021	09/02/2021
									A070909	Oversensing	09/02/2021	09/02/2021
									A0712	Pacing Problem	09/02/2021	09/02/2021
									A072201	High Impedance	09/02/2021	09/02/2021
									A090801	Signal Artifact/Noise	09/02/2021	09/02/2021
							F12	Serious Injury/ Illness/ Impairment	A070101	Failure to Capture	09/02/2021	09/02/2021
									A070909	Oversensing	09/02/2021	09/02/2021
									A0712	Pacing Problem	09/02/2021	09/02/2021
									A072201	High Impedance	09/02/2021	09/02/2021
									A090801	Signal Artifact/Noise	09/02/2021	09/02/2021
							F15	Recognized Device or Procedural Complication	A070101	Failure to Capture	09/02/2021	09/02/2021
									A070909	Oversensing	09/02/2021	09/02/2021
									A0712	Pacing Problem	09/02/2021	09/02/2021
									A072201	High Impedance	09/02/2021	09/02/2021
									A090801	Signal Artifact/Noise	09/02/2021	09/02/2021
							F1903	Device Explantation	A070101	Failure to Capture	09/02/2021	09/02/2021
									A070909	Oversensing	09/02/2021	09/02/2021
									A0712	Pacing Problem	09/02/2021	09/02/2021
									A072201	High Impedance	09/02/2021	09/02/2021
									A090801	Signal Artifact/Noise	09/02/2021	09/02/2021
						CT257	Implantable cardiac pacemakers	E011903	Syncope/Fainting	A070101	Failure to Capture	09/02/2021	09/02/2021
										A070909	Oversensing	09/02/2021	09/02/2021
										A0712	Pacing Problem	09/02/2021	09/02/2021
										A072201	High Impedance	09/02/2021	09/02/2021
										A090801	Signal Artifact/Noise	09/02/2021	09/02/2021
								E060104	Bradycardia	A070101	Failure to Capture	09/02/2021	09/02/2021
										A070909	Oversensing	09/02/2021	09/02/2021
										A0712	Pacing Problem	09/02/2021	09/02/2021
										A072201	High Impedance	09/02/2021	09/02/2021
										A090801	Signal Artifact/Noise	09/02/2021	09/02/2021
								F08	Hospitalization or Prolonged Hospitalization	A070101	Failure to Capture	09/02/2021	09/02/2021
										A070909	Oversensing	09/02/2021	09/02/2021
										A0712	Pacing Problem	09/02/2021	09/02/2021
										A072201	High Impedance	09/02/2021	09/02/2021
										A090801	Signal Artifact/Noise	09/02/2021	09/02/2021
								F12	Serious Injury/ Illness/ Impairment	A070101	Failure to Capture	09/02/2021	09/02/2021
										A070909	Oversensing	09/02/2021	09/02/2021
										A0712	Pacing Problem	09/02/2021	09/02/2021
										A072201	High Impedance	09/02/2021	09/02/2021
										A090801	Signal Artifact/Noise	09/02/2021	09/02/2021
								F15	Recognized Device or Procedural Complication	A070101	Failure to Capture	09/02/2021	09/02/2021
										A070909	Oversensing	09/02/2021	09/02/2021
										A0712	Pacing Problem	09/02/2021	09/02/2021
										A072201	High Impedance	09/02/2021	09/02/2021
										A090801	Signal Artifact/Noise	09/02/2021	09/02/2021
								F1903	Device Explantation	A070101	Failure to Capture	09/02/2021	09/02/2021
										A070909	Oversensing	09/02/2021	09/02/2021
										A0712	Pacing Problem	09/02/2021	09/02/2021
										A072201	High Impedance	09/02/2021	09/02/2021
										A090801	Signal Artifact/Noise	09/02/2021	09/02/2021
						CT2727	Active implant devices/ancillaries	E011903	Syncope/Fainting	A070101	Failure to Capture	09/02/2021	09/02/2021
										A070909	Oversensing	09/02/2021	09/02/2021
										A0712	Pacing Problem	09/02/2021	09/02/2021
										A072201	High Impedance	09/02/2021	09/02/2021
										A090801	Signal Artifact/Noise	09/02/2021	09/02/2021
								E060104	Bradycardia	A070101	Failure to Capture	09/02/2021	09/02/2021
										A070909	Oversensing	09/02/2021	09/02/2021
										A0712	Pacing Problem	09/02/2021	09/02/2021
										A072201	High Impedance	09/02/2021	09/02/2021
										A090801	Signal Artifact/Noise	09/02/2021	09/02/2021
							F08	Hospitalization or Prolonged Hospitalization	A070101	Failure to Capture	09/02/2021	09/02/2021
									A070909	Oversensing	09/02/2021	09/02/2021
									A0712	Pacing Problem	09/02/2021	09/02/2021
									A072201	High Impedance	09/02/2021	09/02/2021
									A090801	Signal Artifact/Noise	09/02/2021	09/02/2021
							F12	Serious Injury/ Illness/ Impairment	A070101	Failure to Capture	09/02/2021	09/02/2021
									A070909	Oversensing	09/02/2021	09/02/2021
									A0712	Pacing Problem	09/02/2021	09/02/2021
									A072201	High Impedance	09/02/2021	09/02/2021
									A090801	Signal Artifact/Noise	09/02/2021	09/02/2021
							F15	Recognized Device or Procedural Complication	A070101	Failure to Capture	09/02/2021	09/02/2021
									A070909	Oversensing	09/02/2021	09/02/2021
									A0712	Pacing Problem	09/02/2021	09/02/2021
									A072201	High Impedance	09/02/2021	09/02/2021
									A090801	Signal Artifact/Noise	09/02/2021	09/02/2021
							F1903	Device Explantation	A070101	Failure to Capture	09/02/2021	09/02/2021
									A070909	Oversensing	09/02/2021	09/02/2021
									A0712	Pacing Problem	09/02/2021	09/02/2021
									A072201	High Impedance	09/02/2021	09/02/2021
									A090801	Signal Artifact/Noise	09/02/2021	09/02/2021
						CT2750	Implantable cardiac electrical stimulators and associated devices	E011903	Syncope/Fainting	A070101	Failure to Capture	09/02/2021	09/02/2021
										A070909	Oversensing	09/02/2021	09/02/2021
										A0712	Pacing Problem	09/02/2021	09/02/2021
										A072201	High Impedance	09/02/2021	09/02/2021
										A090801	Signal Artifact/Noise	09/02/2021	09/02/2021
								E060104	Bradycardia	A070101	Failure to Capture	09/02/2021	09/02/2021
										A070909	Oversensing	09/02/2021	09/02/2021
										A0712	Pacing Problem	09/02/2021	09/02/2021

								A072201	High Impedance	09/02/2021	09/02/2021
								A090801	Signal Artifact/Noise	09/02/2021	09/02/2021
						F08	Hospitalization or Prolonged Hospitalization	A070101	Failure to Capture	09/02/2021	09/02/2021
								A070909	Oversensing	09/02/2021	09/02/2021
								A0712	Pacing Problem	09/02/2021	09/02/2021
								A072201	High Impedance	09/02/2021	09/02/2021
								A090801	Signal Artifact/Noise	09/02/2021	09/02/2021
						F12	Serious Injury/ Illness/ Impairment	A070101	Failure to Capture	09/02/2021	09/02/2021
								A070909	Oversensing	09/02/2021	09/02/2021
								A0712	Pacing Problem	09/02/2021	09/02/2021
								A072201	High Impedance	09/02/2021	09/02/2021
								A090801	Signal Artifact/Noise	09/02/2021	09/02/2021
						F15	Recognized Device or Procedural Complication	A070101	Failure to Capture	09/02/2021	09/02/2021
								A070909	Oversensing	09/02/2021	09/02/2021
								A0712	Pacing Problem	09/02/2021	09/02/2021
								A072201	High Impedance	09/02/2021	09/02/2021
								A090801	Signal Artifact/Noise	09/02/2021	09/02/2021
						F1903	Device Explantation	A070101	Failure to Capture	09/02/2021	09/02/2021
								A070909	Oversensing	09/02/2021	09/02/2021
								A0712	Pacing Problem	09/02/2021	09/02/2021
								A072201	High Impedance	09/02/2021	09/02/2021
								A090801	Signal Artifact/Noise	09/02/2021	09/02/2021
					CT2866	Implantable cardiac electrical stimulators	E011903	Syncope/Fainting	A070101	Failure to Capture	09/02/2021	09/02/2021
									A070909	Oversensing	09/02/2021	09/02/2021
									A0712	Pacing Problem	09/02/2021	09/02/2021
									A072201	High Impedance	09/02/2021	09/02/2021
									A090801	Signal Artifact/Noise	09/02/2021	09/02/2021
							E060104	Bradycardia	A070101	Failure to Capture	09/02/2021	09/02/2021
									A070909	Oversensing	09/02/2021	09/02/2021
									A0712	Pacing Problem	09/02/2021	09/02/2021
									A072201	High Impedance	09/02/2021	09/02/2021
									A090801	Signal Artifact/Noise	09/02/2021	09/02/2021
							F08	Hospitalization or Prolonged Hospitalization	A070101	Failure to Capture	09/02/2021	09/02/2021
									A070909	Oversensing	09/02/2021	09/02/2021
									A0712	Pacing Problem	09/02/2021	09/02/2021
									A072201	High Impedance	09/02/2021	09/02/2021
									A090801	Signal Artifact/Noise	09/02/2021	09/02/2021
							F12	Serious Injury/ Illness/ Impairment	A070101	Failure to Capture	09/02/2021	09/02/2021
									A070909	Oversensing	09/02/2021	09/02/2021
									A0712	Pacing Problem	09/02/2021	09/02/2021
									A072201	High Impedance	09/02/2021	09/02/2021
									A090801	Signal Artifact/Noise	09/02/2021	09/02/2021
							F15	Recognized Device or Procedural Complication	A070101	Failure to Capture	09/02/2021	09/02/2021
									A070909	Oversensing	09/02/2021	09/02/2021
									A0712	Pacing Problem	09/02/2021	09/02/2021
									A072201	High Impedance	09/02/2021	09/02/2021
									A090801	Signal Artifact/Noise	09/02/2021	09/02/2021
							F1903	Device Explantation	A070101	Failure to Capture	09/02/2021	09/02/2021
									A070909	Oversensing	09/02/2021	09/02/2021
									A0712	Pacing Problem	09/02/2021	09/02/2021
									A072201	High Impedance	09/02/2021	09/02/2021
									A090801	Signal Artifact/Noise	09/02/2021	09/02/2021
					CT2985	Active	E011903	Syncope/Fainting	A070101	Failure to Capture	09/02/2021	09/02/2021
									A070909	Oversensing	09/02/2021	09/02/2021
									A0712	Pacing Problem	09/02/2021	09/02/2021
									A072201	High Impedance	09/02/2021	09/02/2021
									A090801	Signal Artifact/Noise	09/02/2021	09/02/2021
							E060104	Bradycardia	A070101	Failure to Capture	09/02/2021	09/02/2021
									A070909	Oversensing	09/02/2021	09/02/2021
									A0712	Pacing Problem	09/02/2021	09/02/2021
									A072201	High Impedance	09/02/2021	09/02/2021
									A090801	Signal Artifact/Noise	09/02/2021	09/02/2021
							F08	Hospitalization or Prolonged Hospitalization	A070101	Failure to Capture	09/02/2021	09/02/2021
									A070909	Oversensing	09/02/2021	09/02/2021
									A0712	Pacing Problem	09/02/2021	09/02/2021
									A072201	High Impedance	09/02/2021	09/02/2021
									A090801	Signal Artifact/Noise	09/02/2021	09/02/2021
							F12	Serious Injury/ Illness/ Impairment	A070101	Failure to Capture	09/02/2021	09/02/2021
									A070909	Oversensing	09/02/2021	09/02/2021
									A0712	Pacing Problem	09/02/2021	09/02/2021
									A072201	High Impedance	09/02/2021	09/02/2021
									A090801	Signal Artifact/Noise	09/02/2021	09/02/2021
							F15	Recognized Device or Procedural Complication	A070101	Failure to Capture	09/02/2021	09/02/2021
									A070909	Oversensing	09/02/2021	09/02/2021
									A0712	Pacing Problem	09/02/2021	09/02/2021
									A072201	High Impedance	09/02/2021	09/02/2021
									A090801	Signal Artifact/Noise	09/02/2021	09/02/2021
							F1903	Device Explantation	A070101	Failure to Capture	09/02/2021	09/02/2021
									A070909	Oversensing	09/02/2021	09/02/2021
									A0712	Pacing Problem	09/02/2021	09/02/2021
									A072201	High Impedance	09/02/2021	09/02/2021
									A090801	Signal Artifact/Noise	09/02/2021	09/02/2021
					CT3001	Implantable metal	E011903	Syncope/Fainting	A070101	Failure to Capture	09/02/2021	09/02/2021
									A070909	Oversensing	09/02/2021	09/02/2021
									A0712	Pacing Problem	09/02/2021	09/02/2021
									A072201	High Impedance	09/02/2021	09/02/2021
									A090801	Signal Artifact/Noise	09/02/2021	09/02/2021
							E060104	Bradycardia	A070101	Failure to Capture	09/02/2021	09/02/2021

								A070909	Oversensing	09/02/2021	09/02/2021
								A0712	Pacing Problem	09/02/2021	09/02/2021
								A072201	High Impedance	09/02/2021	09/02/2021
								A090801	Signal Artifact/Noise	09/02/2021	09/02/2021
						F08	Hospitalization or Prolonged Hospitalization	A070101	Failure to Capture	09/02/2021	09/02/2021
								A070909	Oversensing	09/02/2021	09/02/2021
								A0712	Pacing Problem	09/02/2021	09/02/2021
								A072201	High Impedance	09/02/2021	09/02/2021
								A090801	Signal Artifact/Noise	09/02/2021	09/02/2021
						F12	Serious Injury/ Illness/ Impairment	A070101	Failure to Capture	09/02/2021	09/02/2021
								A070909	Oversensing	09/02/2021	09/02/2021
								A0712	Pacing Problem	09/02/2021	09/02/2021
								A072201	High Impedance	09/02/2021	09/02/2021
								A090801	Signal Artifact/Noise	09/02/2021	09/02/2021
						F15	Recognized Device or Procedural Complication	A070101	Failure to Capture	09/02/2021	09/02/2021
								A070909	Oversensing	09/02/2021	09/02/2021
								A0712	Pacing Problem	09/02/2021	09/02/2021
								A072201	High Impedance	09/02/2021	09/02/2021
								A090801	Signal Artifact/Noise	09/02/2021	09/02/2021
						F1903	Device Explantation	A070101	Failure to Capture	09/02/2021	09/02/2021
								A070909	Oversensing	09/02/2021	09/02/2021
								A0712	Pacing Problem	09/02/2021	09/02/2021
								A072201	High Impedance	09/02/2021	09/02/2021
								A090801	Signal Artifact/Noise	09/02/2021	09/02/2021
					CT3002	Implantable synthetic polymer	E011903	Syncope/Fainting	A070101	Failure to Capture	09/02/2021	09/02/2021
									A070909	Oversensing	09/02/2021	09/02/2021
									A0712	Pacing Problem	09/02/2021	09/02/2021
									A072201	High Impedance	09/02/2021	09/02/2021
									A090801	Signal Artifact/Noise	09/02/2021	09/02/2021
							E060104	Bradycardia	A070101	Failure to Capture	09/02/2021	09/02/2021
									A070909	Oversensing	09/02/2021	09/02/2021
									A0712	Pacing Problem	09/02/2021	09/02/2021
									A072201	High Impedance	09/02/2021	09/02/2021
									A090801	Signal Artifact/Noise	09/02/2021	09/02/2021
							F08	Hospitalization or Prolonged Hospitalization	A070101	Failure to Capture	09/02/2021	09/02/2021
									A070909	Oversensing	09/02/2021	09/02/2021
									A0712	Pacing Problem	09/02/2021	09/02/2021
									A072201	High Impedance	09/02/2021	09/02/2021
									A090801	Signal Artifact/Noise	09/02/2021	09/02/2021
							F12	Serious Injury/ Illness/ Impairment	A070101	Failure to Capture	09/02/2021	09/02/2021
									A070909	Oversensing	09/02/2021	09/02/2021
									A0712	Pacing Problem	09/02/2021	09/02/2021
									A072201	High Impedance	09/02/2021	09/02/2021
									A090801	Signal Artifact/Noise	09/02/2021	09/02/2021
							F15	Recognized Device or Procedural Complication	A070101	Failure to Capture	09/02/2021	09/02/2021
									A070909	Oversensing	09/02/2021	09/02/2021
									A0712	Pacing Problem	09/02/2021	09/02/2021
									A072201	High Impedance	09/02/2021	09/02/2021
									A090801	Signal Artifact/Noise	09/02/2021	09/02/2021
							F1903	Device Explantation	A070101	Failure to Capture	09/02/2021	09/02/2021
									A070909	Oversensing	09/02/2021	09/02/2021
									A0712	Pacing Problem	09/02/2021	09/02/2021
									A072201	High Impedance	09/02/2021	09/02/2021
									A090801	Signal Artifact/Noise	09/02/2021	09/02/2021
									A070909	Oversensing	09/02/2021	09/02/2021
									A0712	Pacing Problem	09/02/2021	09/02/2021
									A072201	High Impedance	09/02/2021	09/02/2021
									A090801	Signal Artifact/Noise	09/02/2021	09/02/2021
							E011903	Syncope/Fainting	A070101	Failure to Capture	09/02/2021	09/02/2021
									A070909	Oversensing	09/02/2021	09/02/2021
									A0712	Pacing Problem	09/02/2021	09/02/2021
									A072201	High Impedance	09/02/2021	09/02/2021
									A090801	Signal Artifact/Noise	09/02/2021	09/02/2021
							E060104	Bradycardia	A070101	Failure to Capture	09/02/2021	09/02/2021
									A070909	Oversensing	09/02/2021	09/02/2021
									A0712	Pacing Problem	09/02/2021	09/02/2021
									A072201	High Impedance	09/02/2021	09/02/2021
									A090801	Signal Artifact/Noise	09/02/2021	09/02/2021
							F08	Hospitalization or Prolonged Hospitalization	A070101	Failure to Capture	09/02/2021	09/02/2021
									A070909	Oversensing	09/02/2021	09/02/2021
									A0712	Pacing Problem	09/02/2021	09/02/2021
									A072201	High Impedance	09/02/2021	09/02/2021
									A090801	Signal Artifact/Noise	09/02/2021	09/02/2021
							F12	Serious Injury/ Illness/ Impairment	A070101	Failure to Capture	09/02/2021	09/02/2021
									A070909	Oversensing	09/02/2021	09/02/2021
									A0712	Pacing Problem	09/02/2021	09/02/2021
									A072201	High Impedance	09/02/2021	09/02/2021
									A090801	Signal Artifact/Noise	09/02/2021	09/02/2021
							F15	Recognized Device or Procedural Complication	A070101	Failure to Capture	09/02/2021	09/02/2021
									A070909	Oversensing	09/02/2021	09/02/2021
									A0712	Pacing Problem	09/02/2021	09/02/2021
									A072201	High Impedance	09/02/2021	09/02/2021
									A090801	Signal Artifact/Noise	09/02/2021	09/02/2021
							F1903	Device Explantation	A070101	Failure to Capture	09/02/2021	09/02/2021
									A070909	Oversensing	09/02/2021	09/02/2021
									A0712	Pacing Problem	09/02/2021	09/02/2021
									A072201	High Impedance	09/02/2021	09/02/2021
									A090801	Signal Artifact/Noise	09/02/2021	09/02/2021
					CT314	Electrically powered	E011903	Syncope/Fainting	A070101	Failure to Capture	09/02/2021	09/02/2021
									A070909	Oversensing	09/02/2021	09/02/2021
									A0712	Pacing Problem	09/02/2021	09/02/2021
									A072201	High Impedance	09/02/2021	09/02/2021

								A090801	Signal Artifact/Noise	09/02/2021	09/02/2021
							E060104	Bradycardia	A070101	Failure to Capture	09/02/2021	09/02/2021
									A070909	Oversensing	09/02/2021	09/02/2021
									A0712	Pacing Problem	09/02/2021	09/02/2021
									A072201	High Impedance	09/02/2021	09/02/2021
									A090801	Signal Artifact/Noise	09/02/2021	09/02/2021
							F08	Hospitalization or Prolonged Hospitalization	A070101	Failure to Capture	09/02/2021	09/02/2021
									A070909	Oversensing	09/02/2021	09/02/2021
									A0712	Pacing Problem	09/02/2021	09/02/2021
									A072201	High Impedance	09/02/2021	09/02/2021
									A090801	Signal Artifact/Noise	09/02/2021	09/02/2021
							F12	Serious Injury/ Illness/ Impairment	A070101	Failure to Capture	09/02/2021	09/02/2021
									A070909	Oversensing	09/02/2021	09/02/2021
									A0712	Pacing Problem	09/02/2021	09/02/2021
									A072201	High Impedance	09/02/2021	09/02/2021
									A090801	Signal Artifact/Noise	09/02/2021	09/02/2021
							F15	Recognized Device or Procedural Complication	A070101	Failure to Capture	09/02/2021	09/02/2021
									A070909	Oversensing	09/02/2021	09/02/2021
									A0712	Pacing Problem	09/02/2021	09/02/2021
									A072201	High Impedance	09/02/2021	09/02/2021
									A090801	Signal Artifact/Noise	09/02/2021	09/02/2021
							F1903	Device Explantation	A070101	Failure to Capture	09/02/2021	09/02/2021
									A070909	Oversensing	09/02/2021	09/02/2021
									A0712	Pacing Problem	09/02/2021	09/02/2021
									A072201	High Impedance	09/02/2021	09/02/2021
									A090801	Signal Artifact/Noise	09/02/2021	09/02/2021
						CT321	Long-term surgical invasive	E011903	Syncope/Fainting	A070101	Failure to Capture	09/02/2021	09/02/2021
										A070909	Oversensing	09/02/2021	09/02/2021
										A0712	Pacing Problem	09/02/2021	09/02/2021
										A072201	High Impedance	09/02/2021	09/02/2021
										A090801	Signal Artifact/Noise	09/02/2021	09/02/2021
								E060104	Bradycardia	A070101	Failure to Capture	09/02/2021	09/02/2021
										A070909	Oversensing	09/02/2021	09/02/2021
										A0712	Pacing Problem	09/02/2021	09/02/2021
										A072201	High Impedance	09/02/2021	09/02/2021
										A090801	Signal Artifact/Noise	09/02/2021	09/02/2021
								F08	Hospitalization or Prolonged Hospitalization	A070101	Failure to Capture	09/02/2021	09/02/2021
										A070909	Oversensing	09/02/2021	09/02/2021
										A0712	Pacing Problem	09/02/2021	09/02/2021
										A072201	High Impedance	09/02/2021	09/02/2021
										A090801	Signal Artifact/Noise	09/02/2021	09/02/2021
								F12	Serious Injury/ Illness/ Impairment	A070101	Failure to Capture	09/02/2021	09/02/2021
										A070909	Oversensing	09/02/2021	09/02/2021
										A0712	Pacing Problem	09/02/2021	09/02/2021
										A072201	High Impedance	09/02/2021	09/02/2021
										A090801	Signal Artifact/Noise	09/02/2021	09/02/2021
								F15	Recognized Device or Procedural Complication	A070101	Failure to Capture	09/02/2021	09/02/2021
										A070909	Oversensing	09/02/2021	09/02/2021
										A0712	Pacing Problem	09/02/2021	09/02/2021
										A072201	High Impedance	09/02/2021	09/02/2021
										A090801	Signal Artifact/Noise	09/02/2021	09/02/2021
								F1903	Device Explantation	A070101	Failure to Capture	09/02/2021	09/02/2021
										A070909	Oversensing	09/02/2021	09/02/2021
										A0712	Pacing Problem	09/02/2021	09/02/2021
										A072201	High Impedance	09/02/2021	09/02/2021
										A090801	Signal Artifact/Noise	09/02/2021	09/02/2021
						CT449	Generators and associated devices	E011903	Syncope/Fainting	A070101	Failure to Capture	09/02/2021	09/02/2021
										A070909	Oversensing	09/02/2021	09/02/2021
										A0712	Pacing Problem	09/02/2021	09/02/2021
										A072201	High Impedance	09/02/2021	09/02/2021
										A090801	Signal Artifact/Noise	09/02/2021	09/02/2021
								E060104	Bradycardia	A070101	Failure to Capture	09/02/2021	09/02/2021
										A070909	Oversensing	09/02/2021	09/02/2021
										A0712	Pacing Problem	09/02/2021	09/02/2021
										A072201	High Impedance	09/02/2021	09/02/2021
										A090801	Signal Artifact/Noise	09/02/2021	09/02/2021
								F08	Hospitalization or Prolonged Hospitalization	A070101	Failure to Capture	09/02/2021	09/02/2021
										A070909	Oversensing	09/02/2021	09/02/2021
										A0712	Pacing Problem	09/02/2021	09/02/2021
										A072201	High Impedance	09/02/2021	09/02/2021
										A090801	Signal Artifact/Noise	09/02/2021	09/02/2021
								F12	Serious Injury/ Illness/ Impairment	A070101	Failure to Capture	09/02/2021	09/02/2021
										A070909	Oversensing	09/02/2021	09/02/2021
										A0712	Pacing Problem	09/02/2021	09/02/2021
										A072201	High Impedance	09/02/2021	09/02/2021
										A090801	Signal Artifact/Noise	09/02/2021	09/02/2021
								F15	Recognized Device or Procedural Complication	A070101	Failure to Capture	09/02/2021	09/02/2021
										A070909	Oversensing	09/02/2021	09/02/2021
										A0712	Pacing Problem	09/02/2021	09/02/2021
										A072201	High Impedance	09/02/2021	09/02/2021
										A090801	Signal Artifact/Noise	09/02/2021	09/02/2021
								F1903	Device Explantation	A070101	Failure to Capture	09/02/2021	09/02/2021
										A070909	Oversensing	09/02/2021	09/02/2021
										A0712	Pacing Problem	09/02/2021	09/02/2021
										A072201	High Impedance	09/02/2021	09/02/2021
										A090801	Signal Artifact/Noise	09/02/2021	09/02/2021
						CT981	Single use	E011903	Syncope/Fainting	A070101	Failure to Capture	09/02/2021	09/02/2021
										A070909	Oversensing	09/02/2021	09/02/2021

									A0712	Pacing Problem	09/02/2021	09/02/2021
									A072201	High Impedance	09/02/2021	09/02/2021
									A090801	Signal Artifact/Noise	09/02/2021	09/02/2021
							E060104	Bradycardia	A070101	Failure to Capture	09/02/2021	09/02/2021
									A070909	Oversensing	09/02/2021	09/02/2021
									A0712	Pacing Problem	09/02/2021	09/02/2021
									A072201	High Impedance	09/02/2021	09/02/2021
									A090801	Signal Artifact/Noise	09/02/2021	09/02/2021
							F08	Hospitalization or Prolonged Hospitalization	A070101	Failure to Capture	09/02/2021	09/02/2021
									A070909	Oversensing	09/02/2021	09/02/2021
									A0712	Pacing Problem	09/02/2021	09/02/2021
									A072201	High Impedance	09/02/2021	09/02/2021
									A090801	Signal Artifact/Noise	09/02/2021	09/02/2021
							F12	Serious Injury/ Illness/ Impairment	A070101	Failure to Capture	09/02/2021	09/02/2021
									A070909	Oversensing	09/02/2021	09/02/2021
									A0712	Pacing Problem	09/02/2021	09/02/2021
									A072201	High Impedance	09/02/2021	09/02/2021
									A090801	Signal Artifact/Noise	09/02/2021	09/02/2021
							F15	Recognized Device or Procedural Complication	A070101	Failure to Capture	09/02/2021	09/02/2021
									A070909	Oversensing	09/02/2021	09/02/2021
									A0712	Pacing Problem	09/02/2021	09/02/2021
									A072201	High Impedance	09/02/2021	09/02/2021
									A090801	Signal Artifact/Noise	09/02/2021	09/02/2021
							F1903	Device Explantation	A070101	Failure to Capture	09/02/2021	09/02/2021
									A070909	Oversensing	09/02/2021	09/02/2021
									A0712	Pacing Problem	09/02/2021	09/02/2021
									A072201	High Impedance	09/02/2021	09/02/2021
									A090801	Signal Artifact/Noise	09/02/2021	09/02/2021
					CT983	Surgical invasive	E011903	Syncope/Fainting	A070101	Failure to Capture	09/02/2021	09/02/2021
									A070909	Oversensing	09/02/2021	09/02/2021
									A0712	Pacing Problem	09/02/2021	09/02/2021
									A072201	High Impedance	09/02/2021	09/02/2021
									A090801	Signal Artifact/Noise	09/02/2021	09/02/2021
							E060104	Bradycardia	A070101	Failure to Capture	09/02/2021	09/02/2021
									A070909	Oversensing	09/02/2021	09/02/2021
									A0712	Pacing Problem	09/02/2021	09/02/2021
									A072201	High Impedance	09/02/2021	09/02/2021
									A090801	Signal Artifact/Noise	09/02/2021	09/02/2021
							F08	Hospitalization or Prolonged Hospitalization	A070101	Failure to Capture	09/02/2021	09/02/2021
									A070909	Oversensing	09/02/2021	09/02/2021
									A0712	Pacing Problem	09/02/2021	09/02/2021
									A072201	High Impedance	09/02/2021	09/02/2021
									A090801	Signal Artifact/Noise	09/02/2021	09/02/2021
							F12	Serious Injury/ Illness/ Impairment	A070101	Failure to Capture	09/02/2021	09/02/2021
									A070909	Oversensing	09/02/2021	09/02/2021
									A0712	Pacing Problem	09/02/2021	09/02/2021
									A072201	High Impedance	09/02/2021	09/02/2021
									A090801	Signal Artifact/Noise	09/02/2021	09/02/2021
							F15	Recognized Device or Procedural Complication	A070101	Failure to Capture	09/02/2021	09/02/2021
									A070909	Oversensing	09/02/2021	09/02/2021
									A0712	Pacing Problem	09/02/2021	09/02/2021
									A072201	High Impedance	09/02/2021	09/02/2021
									A090801	Signal Artifact/Noise	09/02/2021	09/02/2021
							F1903	Device Explantation	A070101	Failure to Capture	09/02/2021	09/02/2021
									A070909	Oversensing	09/02/2021	09/02/2021
									A0712	Pacing Problem	09/02/2021	09/02/2021
									A072201	High Impedance	09/02/2021	09/02/2021
									A090801	Signal Artifact/Noise	09/02/2021	09/02/2021
10582	2021	All other reportable incidents	35852	Single-chamber implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem	24/02/2021	24/02/2021
									A070504	Premature Discharge of Battery	24/02/2021	24/02/2021
							F19	Surgical Intervention	A0705	Battery Problem	24/02/2021	24/02/2021
									A070504	Premature Discharge of Battery	24/02/2021	24/02/2021
10583	2021	All other reportable incidents	17882	Physiologic-monitoring defibrillation system	CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070803	Failure to Power Up		
							F27	Problem Identified During Non-Clinical Procedure	A070803	Failure to Power Up		
10584	2021	All other reportable incidents	37265	Dual-chamber implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	01/02/2021	01/02/2021
							F1901	Additional Surgery	A070504	Premature Discharge of Battery	01/02/2021	01/02/2021
10585	2021	All other reportable incidents	37265	Dual-chamber implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A070912	Incorrect Interpretation of Signal		
									A071302	Inappropriate/Inadequate Shock/Stimulation		
							F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A070912	Incorrect Interpretation of Signal		
									A071302	Inappropriate/Inadequate Shock/Stimulation		
							F23	Unexpected Medical Intervention	A070912	Incorrect Interpretation of Signal		
									A071302	Inappropriate/Inadequate Shock/Stimulation		
10586	2021	All other reportable incidents	12913	Pacemaker, implantable	CT145	Cardiovascular devices	E2403	No Clinical Signs, Symptoms or Conditions	A070803	Failure to Power Up		
					CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070803	Failure to Power Up		
					CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A070803	Failure to Power Up		
					CT2406	Implantable	E2403	No Clinical Signs, Symptoms or Conditions	A070803	Failure to Power Up		
					CT2407	Active implantable	E2403	No Clinical Signs, Symptoms or Conditions	A070803	Failure to Power Up		
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A070803	Failure to Power Up		
					CT2727	Active implant devices/ancillaries	E2403	No Clinical Signs, Symptoms or Conditions	A070803	Failure to Power Up		
					CT2750	Implantable cardiac electrical stimulators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070803	Failure to Power Up		
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A070803	Failure to Power Up		
					CT2985	Active	E2403	No Clinical Signs, Symptoms or Conditions	A070803	Failure to Power Up		
					CT314	Electrically powered	E2403	No Clinical Signs, Symptoms or Conditions	A070803	Failure to Power Up		
10587	2021	All other reportable incidents	12913	Pacemaker, implantable	CT145	Cardiovascular devices	F1905	Device Revision or Replacement	A070803	Failure to Power Up		
					CT1684	Cardiac pacemakers and associated devices	F1905	Device Revision or Replacement	A070803	Failure to Power Up		
					CT1685	Cardiac pacemakers	F1905	Device Revision or Replacement	A070803	Failure to Power Up		
					CT2406	Implantable	F1905	Device Revision or Replacement	A070803	Failure to Power Up		

					CT2407	Active implantable	F1905	Device Revision or Replacement	A070803	Failure to Power Up		
					CT257	Implantable cardiac pacemakers	F1905	Device Revision or Replacement	A070803	Failure to Power Up		
					CT2727	Active implant devices/ancillaries	F1905	Device Revision or Replacement	A070803	Failure to Power Up		
					CT2750	Implantable cardiac electrical stimulators and associated devices	F1905	Device Revision or Replacement	A070803	Failure to Power Up		
					CT2866	Implantable cardiac electrical stimulators	F1905	Device Revision or Replacement	A070803	Failure to Power Up		
					CT2985	Active	F1905	Device Revision or Replacement	A070803	Failure to Power Up		
					CT314	Electrically powered	F1905	Device Revision or Replacement	A070803	Failure to Power Up		
10588	2021	All other reportable incidents	12913	Pacemaker, implantable	CT145	Cardiovascular devices	F1905	Device Revision or Replacement	A070803	Failure to Power Up		
					CT1684	Cardiac pacemakers and associated devices	F1905	Device Revision or Replacement	A070803	Failure to Power Up		
					CT1685	Cardiac pacemakers	F1905	Device Revision or Replacement	A070803	Failure to Power Up		
					CT2406	Implantable	F1905	Device Revision or Replacement	A070803	Failure to Power Up		
					CT2407	Active implantable	F1905	Device Revision or Replacement	A070803	Failure to Power Up		
					CT257	Implantable cardiac pacemakers	F1905	Device Revision or Replacement	A070803	Failure to Power Up		
					CT2727	Active implant devices/ancillaries	F1905	Device Revision or Replacement	A070803	Failure to Power Up		
					CT2750	Implantable cardiac electrical stimulators and associated devices	F1905	Device Revision or Replacement	A070803	Failure to Power Up		
					CT2866	Implantable cardiac electrical stimulators	F1905	Device Revision or Replacement	A070803	Failure to Power Up		
					CT2985	Active	F1905	Device Revision or Replacement	A070803	Failure to Power Up		
					CT314	Electrically powered	F1905	Device Revision or Replacement	A070803	Failure to Power Up		
10589	2021	All other reportable incidents	35223	Endocardial/interventricular septal pacing lead	CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A0401	Break	16/03/2021	16/03/2021
									A040101	Fracture	16/03/2021	16/03/2021
									A070909	Oversensing	16/03/2021	16/03/2021
									A0722	Impedance Problem	16/03/2021	16/03/2021
							F26	No Health Consequences or Impact	A0401	Break	16/03/2021	16/03/2021
									A040101	Fracture	16/03/2021	16/03/2021
									A070909	Oversensing	16/03/2021	16/03/2021
									A0722	Impedance Problem	16/03/2021	16/03/2021
10590	2021	All other reportable incidents	35223	Endocardial/interventricular septal pacing lead	CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070909	Oversensing		
									A090801	Signal Artifact/Noise		
							F1905	Device Revision or Replacement	A070909	Oversensing		
									A090801	Signal Artifact/Noise		
							F26	No Health Consequences or Impact	A070909	Oversensing		
									A090801	Signal Artifact/Noise		
10591	2021	All other reportable incidents	47265	Dual-chamber implantable pacemaker, rate-responsive	CT141	Electrodes/Leads and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	28/03/2021	28/03/2021
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	28/03/2021	28/03/2021
							F12	Serious Injury/ Illness/ Impairment	A070504	Premature Discharge of Battery	28/03/2021	28/03/2021
							F19	Surgical Intervention	A070504	Premature Discharge of Battery	28/03/2021	28/03/2021
					CT145	Cardiovascular devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	28/03/2021	28/03/2021
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	28/03/2021	28/03/2021
							F12	Serious Injury/ Illness/ Impairment	A070504	Premature Discharge of Battery	28/03/2021	28/03/2021
							F19	Surgical Intervention	A070504	Premature Discharge of Battery	28/03/2021	28/03/2021
					CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	28/03/2021	28/03/2021
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	28/03/2021	28/03/2021
							F12	Serious Injury/ Illness/ Impairment	A070504	Premature Discharge of Battery	28/03/2021	28/03/2021
							F19	Surgical Intervention	A070504	Premature Discharge of Battery	28/03/2021	28/03/2021
					CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	28/03/2021	28/03/2021
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	28/03/2021	28/03/2021
							F12	Serious Injury/ Illness/ Impairment	A070504	Premature Discharge of Battery	28/03/2021	28/03/2021
							F19	Surgical Intervention	A070504	Premature Discharge of Battery	28/03/2021	28/03/2021
					CT233	Surgical	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	28/03/2021	28/03/2021
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	28/03/2021	28/03/2021
							F12	Serious Injury/ Illness/ Impairment	A070504	Premature Discharge of Battery	28/03/2021	28/03/2021
							F19	Surgical Intervention	A070504	Premature Discharge of Battery	28/03/2021	28/03/2021
					CT2406	Implantable	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	28/03/2021	28/03/2021
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	28/03/2021	28/03/2021
							F12	Serious Injury/ Illness/ Impairment	A070504	Premature Discharge of Battery	28/03/2021	28/03/2021
							F19	Surgical Intervention	A070504	Premature Discharge of Battery	28/03/2021	28/03/2021
					CT2407	Active implantable	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	28/03/2021	28/03/2021
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	28/03/2021	28/03/2021
							F12	Serious Injury/ Illness/ Impairment	A070504	Premature Discharge of Battery	28/03/2021	28/03/2021
							F19	Surgical Intervention	A070504	Premature Discharge of Battery	28/03/2021	28/03/2021
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	28/03/2021	28/03/2021
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	28/03/2021	28/03/2021
							F12	Serious Injury/ Illness/ Impairment	A070504	Premature Discharge of Battery	28/03/2021	28/03/2021
							F19	Surgical Intervention	A070504	Premature Discharge of Battery	28/03/2021	28/03/2021
					CT2727	Active implant devices/ancillaries	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	28/03/2021	28/03/2021
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	28/03/2021	28/03/2021
							F12	Serious Injury/ Illness/ Impairment	A070504	Premature Discharge of Battery	28/03/2021	28/03/2021
							F19	Surgical Intervention	A070504	Premature Discharge of Battery	28/03/2021	28/03/2021
					CT2750	Implantable cardiac electrical stimulators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	28/03/2021	28/03/2021
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	28/03/2021	28/03/2021
							F12	Serious Injury/ Illness/ Impairment	A070504	Premature Discharge of Battery	28/03/2021	28/03/2021
							F19	Surgical Intervention	A070504	Premature Discharge of Battery	28/03/2021	28/03/2021
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	28/03/2021	28/03/2021
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	28/03/2021	28/03/2021
							F12	Serious Injury/ Illness/ Impairment	A070504	Premature Discharge of Battery	28/03/2021	28/03/2021
							F19	Surgical Intervention	A070504	Premature Discharge of Battery	28/03/2021	28/03/2021
					CT2985	Active	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	28/03/2021	28/03/2021
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	28/03/2021	28/03/2021
							F12	Serious Injury/ Illness/ Impairment	A070504	Premature Discharge of Battery	28/03/2021	28/03/2021
							F19	Surgical Intervention	A070504	Premature Discharge of Battery	28/03/2021	28/03/2021
					CT3001	Implantable metal	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	28/03/2021	28/03/2021
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	28/03/2021	28/03/2021
							F12	Serious Injury/ Illness/ Impairment	A070504	Premature Discharge of Battery	28/03/2021	28/03/2021
							F19	Surgical Intervention	A070504	Premature Discharge of Battery	28/03/2021	28/03/2021
					CT3002	Implantable synthetic polymer	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	28/03/2021	28/03/2021
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	28/03/2021	28/03/2021

						F12	Serious Injury/ Illness/ Impairment	A070504	Premature Discharge of Battery	28/03/2021	28/03/2021
						F19	Surgical Intervention	A070504	Premature Discharge of Battery	28/03/2021	28/03/2021
					CT3006	Implantable inorganic material	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	28/03/2021	28/03/2021
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	28/03/2021	28/03/2021
							F12	Serious Injury/ Illness/ Impairment	A070504	Premature Discharge of Battery	28/03/2021	28/03/2021
							F19	Surgical Intervention	A070504	Premature Discharge of Battery	28/03/2021	28/03/2021
					CT314	Electrically powered	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	28/03/2021	28/03/2021
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	28/03/2021	28/03/2021
							F12	Serious Injury/ Illness/ Impairment	A070504	Premature Discharge of Battery	28/03/2021	28/03/2021
							F19	Surgical Intervention	A070504	Premature Discharge of Battery	28/03/2021	28/03/2021
					CT321	Long-term surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	28/03/2021	28/03/2021
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	28/03/2021	28/03/2021
							F12	Serious Injury/ Illness/ Impairment	A070504	Premature Discharge of Battery	28/03/2021	28/03/2021
							F19	Surgical Intervention	A070504	Premature Discharge of Battery	28/03/2021	28/03/2021
					CT449	Generators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	28/03/2021	28/03/2021
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	28/03/2021	28/03/2021
							F12	Serious Injury/ Illness/ Impairment	A070504	Premature Discharge of Battery	28/03/2021	28/03/2021
							F19	Surgical Intervention	A070504	Premature Discharge of Battery	28/03/2021	28/03/2021
					CT981	Single use	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	28/03/2021	28/03/2021
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	28/03/2021	28/03/2021
							F12	Serious Injury/ Illness/ Impairment	A070504	Premature Discharge of Battery	28/03/2021	28/03/2021
							F19	Surgical Intervention	A070504	Premature Discharge of Battery	28/03/2021	28/03/2021
					CT983	Surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	28/03/2021	28/03/2021
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	28/03/2021	28/03/2021
							F12	Serious Injury/ Illness/ Impairment	A070504	Premature Discharge of Battery	28/03/2021	28/03/2021
							F19	Surgical Intervention	A070504	Premature Discharge of Battery	28/03/2021	28/03/2021
10592	2021	All other reportable incidents	47267	Single-chamber implantable pacemaker, rate-responsive	CT141	Electrodes/Leads and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
					CT145	Cardiovascular devices	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
					CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
					CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
					CT233	Surgical	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
					CT2406	Implantable	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
					CT2407	Active implantable	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
					CT2727	Active implant devices/ancillaries	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
					CT2750	Implantable cardiac electrical stimulators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
					CT2985	Active	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
					CT3001	Implantable metal	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
					CT3002	Implantable synthetic polymer	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
					CT3006	Implantable inorganic material	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
					CT314	Electrically powered	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		

							F26	No Health Consequences or Impact	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
					CT321	Long-term surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
					CT449	Generators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
					CT981	Single use	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
					CT983	Surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
10593	2021	All other reportable incidents	37265	Dual-chamber implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E2401	Insufficient Information	A072201	High Impedance		
							E2403	No Clinical Signs, Symptoms or Conditions	A072201	High Impedance		
							F24	Insufficient Health Impact Information	A072201	High Impedance		
10594	2021	All other reportable incidents	17882	Physiologic-monitoring defibrillation system	CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070803	Failure to Power Up		
							F27	Problem Identified During Non-Clinical Procedure	A070803	Failure to Power Up		
10595	2021	All other reportable incidents	17882	Physiologic-monitoring defibrillation system	CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070803	Failure to Power Up		
							F26	No Health Consequences or Impact	A070803	Failure to Power Up		
10596	2021	All other reportable incidents	17882	Physiologic-monitoring defibrillation system	CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A0713	Defibrillation/Stimulation Problem		
							F27	Problem Identified During Non-Clinical Procedure	A0713	Defibrillation/Stimulation Problem		
10597	2021	All other reportable incidents	17882	Physiologic-monitoring defibrillation system	CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A0801	Failure to Calibrate		
							F27	Problem Identified During Non-Clinical Procedure	A0801	Failure to Calibrate		
10598	2021	All other reportable incidents	35223	Endocardial/interventricular septal pacing lead	CT1684	Cardiac pacemakers and associated devices	E0623	Unspecified Heart Problem	A070909	Oversensing		
									A0712	Pacing Problem		
							E2403	No Clinical Signs, Symptoms or Conditions	A070909	Oversensing		
									A0712	Pacing Problem		
							F23	Unexpected Medical Intervention	A070909	Oversensing		
									A0712	Pacing Problem		
10599	2021	All other reportable incidents	47265	Dual-chamber implantable pacemaker, rate-responsive	CT141	Electrodes/Leads and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
									A09	Output Problem		
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery		
									A09	Output Problem		
							F12	Serious Injury/ Illness/ Impairment	A070504	Premature Discharge of Battery		
									A09	Output Problem		
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery		
									A09	Output Problem		
							F23	Unexpected Medical Intervention	A070504	Premature Discharge of Battery		
									A09	Output Problem		
					CT145	Cardiovascular devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
									A09	Output Problem		
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery		
									A09	Output Problem		
							F12	Serious Injury/ Illness/ Impairment	A070504	Premature Discharge of Battery		
									A09	Output Problem		
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery		
									A09	Output Problem		
							F23	Unexpected Medical Intervention	A070504	Premature Discharge of Battery		
									A09	Output Problem		
					CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
									A09	Output Problem		
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery		
									A09	Output Problem		
							F12	Serious Injury/ Illness/ Impairment	A070504	Premature Discharge of Battery		
									A09	Output Problem		
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery		
									A09	Output Problem		
							F23	Unexpected Medical Intervention	A070504	Premature Discharge of Battery		
									A09	Output Problem		
					CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
									A09	Output Problem		
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery		
									A09	Output Problem		
							F12	Serious Injury/ Illness/ Impairment	A070504	Premature Discharge of Battery		
									A09	Output Problem		
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery		
									A09	Output Problem		
							F23	Unexpected Medical Intervention	A070504	Premature Discharge of Battery		
									A09	Output Problem		
					CT233	Surgical	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
									A09	Output Problem		
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery		
									A09	Output Problem		
							F12	Serious Injury/ Illness/ Impairment	A070504	Premature Discharge of Battery		
									A09	Output Problem		
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery		
									A09	Output Problem		
							F23	Unexpected Medical Intervention	A070504	Premature Discharge of Battery		
									A09	Output Problem		
					CT2406	Implantable	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
									A09	Output Problem		
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery		

									A09	Output Problem		
							F12	Serious Injury/ Illness/ Impairment	A070504	Premature Discharge of Battery		
									A09	Output Problem		
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery		
									A09	Output Problem		
							F23	Unexpected Medical Intervention	A070504	Premature Discharge of Battery		
									A09	Output Problem		
					CT2407	Active implantable	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
									A09	Output Problem		
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery		
									A09	Output Problem		
							F12	Serious Injury/ Illness/ Impairment	A070504	Premature Discharge of Battery		
									A09	Output Problem		
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery		
									A09	Output Problem		
							F23	Unexpected Medical Intervention	A070504	Premature Discharge of Battery		
									A09	Output Problem		
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
									A09	Output Problem		
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery		
									A09	Output Problem		
							F12	Serious Injury/ Illness/ Impairment	A070504	Premature Discharge of Battery		
									A09	Output Problem		
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery		
									A09	Output Problem		
							F23	Unexpected Medical Intervention	A070504	Premature Discharge of Battery		
									A09	Output Problem		
					CT2727	Active implant devices/ancillaries	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
									A09	Output Problem		
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery		
									A09	Output Problem		
							F12	Serious Injury/ Illness/ Impairment	A070504	Premature Discharge of Battery		
									A09	Output Problem		
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery		
									A09	Output Problem		
							F23	Unexpected Medical Intervention	A070504	Premature Discharge of Battery		
									A09	Output Problem		
					CT2750	Implantable cardiac electrical stimulators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
									A09	Output Problem		
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery		
									A09	Output Problem		
							F12	Serious Injury/ Illness/ Impairment	A070504	Premature Discharge of Battery		
									A09	Output Problem		
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery		
									A09	Output Problem		
							F23	Unexpected Medical Intervention	A070504	Premature Discharge of Battery		
									A09	Output Problem		
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
									A09	Output Problem		
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery		
									A09	Output Problem		
							F12	Serious Injury/ Illness/ Impairment	A070504	Premature Discharge of Battery		
									A09	Output Problem		
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery		
									A09	Output Problem		
							F23	Unexpected Medical Intervention	A070504	Premature Discharge of Battery		
									A09	Output Problem		
					CT2985	Active	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
									A09	Output Problem		
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery		
									A09	Output Problem		
							F12	Serious Injury/ Illness/ Impairment	A070504	Premature Discharge of Battery		
									A09	Output Problem		
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery		
									A09	Output Problem		
							F23	Unexpected Medical Intervention	A070504	Premature Discharge of Battery		
									A09	Output Problem		
					CT3001	Implantable metal	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
									A09	Output Problem		
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery		
									A09	Output Problem		
							F12	Serious Injury/ Illness/ Impairment	A070504	Premature Discharge of Battery		
									A09	Output Problem		
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery		
									A09	Output Problem		
							F23	Unexpected Medical Intervention	A070504	Premature Discharge of Battery		
									A09	Output Problem		
					CT3002	Implantable synthetic polymer	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
									A09	Output Problem		
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery		
									A09	Output Problem		
							F12	Serious Injury/ Illness/ Impairment	A070504	Premature Discharge of Battery		
									A09	Output Problem		
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery		
									A09	Output Problem		
							F23	Unexpected Medical Intervention	A070504	Premature Discharge of Battery		
									A09	Output Problem		
					CT3006	Implantable inorganic material	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		

									A09	Output Problem		
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery		
									A09	Output Problem		
							F12	Serious Injury/ Illness/ Impairment	A070504	Premature Discharge of Battery		
									A09	Output Problem		
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery		
									A09	Output Problem		
							F23	Unexpected Medical Intervention	A070504	Premature Discharge of Battery		
									A09	Output Problem		
					CT314	Electrically powered	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
									A09	Output Problem		
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery		
									A09	Output Problem		
							F12	Serious Injury/ Illness/ Impairment	A070504	Premature Discharge of Battery		
									A09	Output Problem		
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery		
									A09	Output Problem		
							F23	Unexpected Medical Intervention	A070504	Premature Discharge of Battery		
									A09	Output Problem		
					CT321	Long-term surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
									A09	Output Problem		
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery		
									A09	Output Problem		
							F12	Serious Injury/ Illness/ Impairment	A070504	Premature Discharge of Battery		
									A09	Output Problem		
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery		
									A09	Output Problem		
							F23	Unexpected Medical Intervention	A070504	Premature Discharge of Battery		
									A09	Output Problem		
					CT449	Generators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
									A09	Output Problem		
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery		
									A09	Output Problem		
							F12	Serious Injury/ Illness/ Impairment	A070504	Premature Discharge of Battery		
									A09	Output Problem		
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery		
									A09	Output Problem		
							F23	Unexpected Medical Intervention	A070504	Premature Discharge of Battery		
									A09	Output Problem		
					CT981	Single use	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
									A09	Output Problem		
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery		
									A09	Output Problem		
							F12	Serious Injury/ Illness/ Impairment	A070504	Premature Discharge of Battery		
									A09	Output Problem		
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery		
									A09	Output Problem		
							F23	Unexpected Medical Intervention	A070504	Premature Discharge of Battery		
									A09	Output Problem		
					CT983	Surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
									A09	Output Problem		
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery		
									A09	Output Problem		
							F12	Serious Injury/ Illness/ Impairment	A070504	Premature Discharge of Battery		
									A09	Output Problem		
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery		
									A09	Output Problem		
							F23	Unexpected Medical Intervention	A070504	Premature Discharge of Battery		
									A09	Output Problem		
10600	2021	All other reportable incidents	35223	Endocardial/interventricular septal pacing lead	CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A090801	Signal Artifact/Noise	28/01/2021	28/01/2021
							F1905	Device Revision or Replacement	A090801	Signal Artifact/Noise	28/01/2021	28/01/2021
							F26	No Health Consequences or Impact	A090801	Signal Artifact/Noise	28/01/2021	28/01/2021
10601	2021	All other reportable incidents	47263	Cardiac resynchronization therapy implantable pacemaker	CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A150206	Difficult to Insert	09/12/2020	09/12/2020
							F14	Prolonged Episode of Care	A150206	Difficult to Insert	09/12/2020	09/12/2020
							F1908	Prolonged Surgery	A150206	Difficult to Insert	09/12/2020	09/12/2020
					CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A150206	Difficult to Insert	09/12/2020	09/12/2020
							F14	Prolonged Episode of Care	A150206	Difficult to Insert	09/12/2020	09/12/2020
							F1908	Prolonged Surgery	A150206	Difficult to Insert	09/12/2020	09/12/2020
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A150206	Difficult to Insert	09/12/2020	09/12/2020
							F14	Prolonged Episode of Care	A150206	Difficult to Insert	09/12/2020	09/12/2020
							F1908	Prolonged Surgery	A150206	Difficult to Insert	09/12/2020	09/12/2020
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A150206	Difficult to Insert	09/12/2020	09/12/2020
							F14	Prolonged Episode of Care	A150206	Difficult to Insert	09/12/2020	09/12/2020
							F1908	Prolonged Surgery	A150206	Difficult to Insert	09/12/2020	09/12/2020
10602	2021	All other reportable incidents	36241	Pacemaker/defibrillator lead	CT1684	Cardiac pacemakers and associated devices	E2104	Electric Shock	A071302	Inappropriate/Inadequate Shock/Stimulation		
							E2403	No Clinical Signs, Symptoms or Conditions	A071302	Inappropriate/Inadequate Shock/Stimulation		
							F26	No Health Consequences or Impact	A071302	Inappropriate/Inadequate Shock/Stimulation		
10603	2021	All other reportable incidents	35852	Single-chamber implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem	16/03/2021	16/03/2021
									A070504	Premature Discharge of Battery	16/03/2021	16/03/2021
							F19	Surgical Intervention	A0705	Battery Problem	16/03/2021	16/03/2021
									A070504	Premature Discharge of Battery	16/03/2021	16/03/2021
10604	2021	All other reportable incidents	35852	Single-chamber implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
							F24	Insufficient Health Impact Information	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
10605	2021	All other reportable incidents	17882	Physiologic-monitoring defibrillation system	CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070803	Failure to Power Up		
							F27	Problem Identified During Non-Clinical Procedure	A070803	Failure to Power Up		
10606	2021	All other reportable incidents	47265	Dual-chamber implantable pacemaker, rate-	CT141	Electrodes/Leads and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem		

				responsive					A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
					CT145	Cardiovascular devices	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
					CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
					CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
					CT233	Surgical	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
					CT2406	Implantable	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
					CT2407	Active implantable	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
					CT2727	Active implant devices/ancillaries	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
					CT2750	Implantable cardiac electrical stimulators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
					CT2985	Active	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
					CT3001	Implantable metal	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
					CT3002	Implantable synthetic polymer	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
					CT3006	Implantable inorganic material	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
					CT314	Electrically powered	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
					CT321	Long-term surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
					CT449	Generators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
					CT981	Single use	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
					CT983	Surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
10607	2021	All other reportable incidents	47265	Dual-chamber implantable pacemaker, rate-responsive	CT141	Electrodes/Leads and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	16/01/2021	16/01/2021
							F19	Surgical Intervention	A070504	Premature Discharge of Battery	16/01/2021	16/01/2021
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	16/01/2021	16/01/2021
					CT145	Cardiovascular devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	16/01/2021	16/01/2021
							F19	Surgical Intervention	A070504	Premature Discharge of Battery	16/01/2021	16/01/2021
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	16/01/2021	16/01/2021
					CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	16/01/2021	16/01/2021
							F19	Surgical Intervention	A070504	Premature Discharge of Battery	16/01/2021	16/01/2021
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	16/01/2021	16/01/2021

					CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	16/01/2021	16/01/2021
							F19	Surgical Intervention	A070504	Premature Discharge of Battery	16/01/2021	16/01/2021
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	16/01/2021	16/01/2021
					CT233	Surgical	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	16/01/2021	16/01/2021
							F19	Surgical Intervention	A070504	Premature Discharge of Battery	16/01/2021	16/01/2021
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	16/01/2021	16/01/2021
					CT2406	Implantable	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	16/01/2021	16/01/2021
							F19	Surgical Intervention	A070504	Premature Discharge of Battery	16/01/2021	16/01/2021
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	16/01/2021	16/01/2021
					CT2407	Active implantable	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	16/01/2021	16/01/2021
							F19	Surgical Intervention	A070504	Premature Discharge of Battery	16/01/2021	16/01/2021
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	16/01/2021	16/01/2021
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	16/01/2021	16/01/2021
							F19	Surgical Intervention	A070504	Premature Discharge of Battery	16/01/2021	16/01/2021
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	16/01/2021	16/01/2021
					CT2727	Active implant devices/ancillaries	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	16/01/2021	16/01/2021
							F19	Surgical Intervention	A070504	Premature Discharge of Battery	16/01/2021	16/01/2021
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	16/01/2021	16/01/2021
					CT2750	Implantable cardiac electrical stimulators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	16/01/2021	16/01/2021
							F19	Surgical Intervention	A070504	Premature Discharge of Battery	16/01/2021	16/01/2021
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	16/01/2021	16/01/2021
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	16/01/2021	16/01/2021
							F19	Surgical Intervention	A070504	Premature Discharge of Battery	16/01/2021	16/01/2021
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	16/01/2021	16/01/2021
					CT2985	Active	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	16/01/2021	16/01/2021
							F19	Surgical Intervention	A070504	Premature Discharge of Battery	16/01/2021	16/01/2021
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	16/01/2021	16/01/2021
					CT3001	Implantable metal	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	16/01/2021	16/01/2021
							F19	Surgical Intervention	A070504	Premature Discharge of Battery	16/01/2021	16/01/2021
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	16/01/2021	16/01/2021
					CT3002	Implantable synthetic polymer	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	16/01/2021	16/01/2021
							F19	Surgical Intervention	A070504	Premature Discharge of Battery	16/01/2021	16/01/2021
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	16/01/2021	16/01/2021
					CT3006	Implantable inorganic material	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	16/01/2021	16/01/2021
							F19	Surgical Intervention	A070504	Premature Discharge of Battery	16/01/2021	16/01/2021
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	16/01/2021	16/01/2021
					CT314	Electrically powered	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	16/01/2021	16/01/2021
							F19	Surgical Intervention	A070504	Premature Discharge of Battery	16/01/2021	16/01/2021
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	16/01/2021	16/01/2021
					CT321	Long-term surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	16/01/2021	16/01/2021
							F19	Surgical Intervention	A070504	Premature Discharge of Battery	16/01/2021	16/01/2021
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	16/01/2021	16/01/2021
					CT449	Generators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	16/01/2021	16/01/2021
							F19	Surgical Intervention	A070504	Premature Discharge of Battery	16/01/2021	16/01/2021
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	16/01/2021	16/01/2021
					CT981	Single use	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	16/01/2021	16/01/2021
							F19	Surgical Intervention	A070504	Premature Discharge of Battery	16/01/2021	16/01/2021
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	16/01/2021	16/01/2021
					CT983	Surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	16/01/2021	16/01/2021
							F19	Surgical Intervention	A070504	Premature Discharge of Battery	16/01/2021	16/01/2021
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	16/01/2021	16/01/2021
10608	2021	All other reportable incidents	36241	Pacemaker/defibrillator lead	CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A25	No Apparent Adverse Event		
							F08	Hospitalization or Prolonged Hospitalization	A25	No Apparent Adverse Event		
							F15	Recognized Device or Procedural Complication	A25	No Apparent Adverse Event		
							F1905	Device Revision or Replacement	A25	No Apparent Adverse Event		
10609	2021	All other reportable incidents	35223	Endocardial/interventricular septal pacing lead	CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A07	Electrical /Electronic Property Problem		
									A072202	Low Impedance		
							F26	No Health Consequences or Impact	A07	Electrical /Electronic Property Problem		
									A072202	Low Impedance		
10610	2021	All other reportable incidents	35852	Single-chamber implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A0908	Incorrect, Inadequate or Imprecise Result or Readings	23/05/2021	23/05/2021
							F27	Problem Identified During Non-Clinical Procedure	A0908	Incorrect, Inadequate or Imprecise Result or Readings	23/05/2021	23/05/2021
10611	2021	All other reportable incidents	35852	Single-chamber implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E2104	Electric Shock	A070909	Oversensing		
									A071302	Inappropriate/Inadequate Shock/Stimulation		
							F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A070909	Oversensing		
									A071302	Inappropriate/Inadequate Shock/Stimulation		
							F23	Unexpected Medical Intervention	A070909	Oversensing		
									A071302	Inappropriate/Inadequate Shock/Stimulation		
10612	2021	All other reportable incidents	37265	Dual-chamber implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E2401	Insufficient Information	A070909	Oversensing		
									A0712	Pacing Problem		
							E2403	No Clinical Signs, Symptoms or Conditions	A070909	Oversensing		
									A0712	Pacing Problem		
							F0101	Therapeutic Response Decreased	A070909	Oversensing		
									A0712	Pacing Problem		
10613	2021	All other reportable incidents	47267	Single-chamber implantable pacemaker, rate-responsive	CT141	Electrodes/Leads and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	09/03/2021	09/03/2021
							F12	Serious Injury/ Illness/ Impairment	A070504	Premature Discharge of Battery	09/03/2021	09/03/2021
							F19	Surgical Intervention	A070504	Premature Discharge of Battery	09/03/2021	09/03/2021
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	09/03/2021	09/03/2021
					CT145	Cardiovascular devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	09/03/2021	09/03/2021
							F12	Serious Injury/ Illness/ Impairment	A070504	Premature Discharge of Battery	09/03/2021	09/03/2021
							F19	Surgical Intervention	A070504	Premature Discharge of Battery	09/03/2021	09/03/2021
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	09/03/2021	09/03/2021
					CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	09/03/2021	09/03/2021
							F12	Serious Injury/ Illness/ Impairment	A070504	Premature Discharge of Battery	09/03/2021	09/03/2021
							F19	Surgical Intervention	A070504	Premature Discharge of Battery	09/03/2021	09/03/2021
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	09/03/2021	09/03/2021
					CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	09/03/2021	09/03/2021

							F12	Serious Injury/ Illness/ Impairment	A070504	Premature Discharge of Battery	09/03/2021	09/03/2021
							F19	Surgical Intervention	A070504	Premature Discharge of Battery	09/03/2021	09/03/2021
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	09/03/2021	09/03/2021
					CT233	Surgical	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	09/03/2021	09/03/2021
							F12	Serious Injury/ Illness/ Impairment	A070504	Premature Discharge of Battery	09/03/2021	09/03/2021
							F19	Surgical Intervention	A070504	Premature Discharge of Battery	09/03/2021	09/03/2021
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	09/03/2021	09/03/2021
					CT2406	Implantable	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	09/03/2021	09/03/2021
							F12	Serious Injury/ Illness/ Impairment	A070504	Premature Discharge of Battery	09/03/2021	09/03/2021
							F19	Surgical Intervention	A070504	Premature Discharge of Battery	09/03/2021	09/03/2021
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	09/03/2021	09/03/2021
					CT2407	Active implantable	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	09/03/2021	09/03/2021
							F12	Serious Injury/ Illness/ Impairment	A070504	Premature Discharge of Battery	09/03/2021	09/03/2021
							F19	Surgical Intervention	A070504	Premature Discharge of Battery	09/03/2021	09/03/2021
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	09/03/2021	09/03/2021
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	09/03/2021	09/03/2021
							F12	Serious Injury/ Illness/ Impairment	A070504	Premature Discharge of Battery	09/03/2021	09/03/2021
							F19	Surgical Intervention	A070504	Premature Discharge of Battery	09/03/2021	09/03/2021
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	09/03/2021	09/03/2021
					CT2727	Active implant devices/ancillaries	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	09/03/2021	09/03/2021
							F12	Serious Injury/ Illness/ Impairment	A070504	Premature Discharge of Battery	09/03/2021	09/03/2021
							F19	Surgical Intervention	A070504	Premature Discharge of Battery	09/03/2021	09/03/2021
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	09/03/2021	09/03/2021
					CT2750	Implantable cardiac electrical stimulators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	09/03/2021	09/03/2021
							F12	Serious Injury/ Illness/ Impairment	A070504	Premature Discharge of Battery	09/03/2021	09/03/2021
							F19	Surgical Intervention	A070504	Premature Discharge of Battery	09/03/2021	09/03/2021
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	09/03/2021	09/03/2021
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	09/03/2021	09/03/2021
							F12	Serious Injury/ Illness/ Impairment	A070504	Premature Discharge of Battery	09/03/2021	09/03/2021
							F19	Surgical Intervention	A070504	Premature Discharge of Battery	09/03/2021	09/03/2021
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	09/03/2021	09/03/2021
					CT2985	Active	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	09/03/2021	09/03/2021
							F12	Serious Injury/ Illness/ Impairment	A070504	Premature Discharge of Battery	09/03/2021	09/03/2021
							F19	Surgical Intervention	A070504	Premature Discharge of Battery	09/03/2021	09/03/2021
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	09/03/2021	09/03/2021
					CT3001	Implantable metal	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	09/03/2021	09/03/2021
							F12	Serious Injury/ Illness/ Impairment	A070504	Premature Discharge of Battery	09/03/2021	09/03/2021
							F19	Surgical Intervention	A070504	Premature Discharge of Battery	09/03/2021	09/03/2021
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	09/03/2021	09/03/2021
					CT3002	Implantable synthetic polymer	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	09/03/2021	09/03/2021
							F12	Serious Injury/ Illness/ Impairment	A070504	Premature Discharge of Battery	09/03/2021	09/03/2021
							F19	Surgical Intervention	A070504	Premature Discharge of Battery	09/03/2021	09/03/2021
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	09/03/2021	09/03/2021
					CT3006	Implantable inorganic material	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	09/03/2021	09/03/2021
							F12	Serious Injury/ Illness/ Impairment	A070504	Premature Discharge of Battery	09/03/2021	09/03/2021
							F19	Surgical Intervention	A070504	Premature Discharge of Battery	09/03/2021	09/03/2021
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	09/03/2021	09/03/2021
					CT314	Electrically powered	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	09/03/2021	09/03/2021
							F12	Serious Injury/ Illness/ Impairment	A070504	Premature Discharge of Battery	09/03/2021	09/03/2021
							F19	Surgical Intervention	A070504	Premature Discharge of Battery	09/03/2021	09/03/2021
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	09/03/2021	09/03/2021
					CT321	Long-term surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	09/03/2021	09/03/2021
							F12	Serious Injury/ Illness/ Impairment	A070504	Premature Discharge of Battery	09/03/2021	09/03/2021
							F19	Surgical Intervention	A070504	Premature Discharge of Battery	09/03/2021	09/03/2021
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	09/03/2021	09/03/2021
					CT449	Generators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	09/03/2021	09/03/2021
							F12	Serious Injury/ Illness/ Impairment	A070504	Premature Discharge of Battery	09/03/2021	09/03/2021
							F19	Surgical Intervention	A070504	Premature Discharge of Battery	09/03/2021	09/03/2021
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	09/03/2021	09/03/2021
					CT981	Single use	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	09/03/2021	09/03/2021
							F12	Serious Injury/ Illness/ Impairment	A070504	Premature Discharge of Battery	09/03/2021	09/03/2021
							F19	Surgical Intervention	A070504	Premature Discharge of Battery	09/03/2021	09/03/2021
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	09/03/2021	09/03/2021
					CT983	Surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	09/03/2021	09/03/2021
							F12	Serious Injury/ Illness/ Impairment	A070504	Premature Discharge of Battery	09/03/2021	09/03/2021
							F19	Surgical Intervention	A070504	Premature Discharge of Battery	09/03/2021	09/03/2021
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	09/03/2021	09/03/2021
10614	2021	All other reportable incidents	37265	Dual-chamber implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A072202	Low Impedance		
									A1301	Failure to Read Input Signal		
							F26	No Health Consequences or Impact	A072202	Low Impedance		
									A1301	Failure to Read Input Signal		
10615	2021	All other reportable incidents	35852	Single-chamber implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E2104	Electric Shock	A071302	Inappropriate/Inadequate Shock/Stimulation		
							E2403	No Clinical Signs, Symptoms or Conditions	A071302	Inappropriate/Inadequate Shock/Stimulation		
							F26	No Health Consequences or Impact	A071302	Inappropriate/Inadequate Shock/Stimulation		
10616	2021	All other reportable incidents	36241	Pacemaker/defibrillator lead	CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A0203	Defective Device		
									A072202	Low Impedance		
							F26	No Health Consequences or Impact	A0203	Defective Device		
									A072202	Low Impedance		
10617	2021	All other reportable incidents	47267	Single-chamber implantable pacemaker, rate-responsive	CT141	Electrodes/Leads and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A25	No Apparent Adverse Event		
							F26	No Health Consequences or Impact	A25	No Apparent Adverse Event		
					CT145	Cardiovascular devices	E2403	No Clinical Signs, Symptoms or Conditions	A25	No Apparent Adverse Event		
							F26	No Health Consequences or Impact	A25	No Apparent Adverse Event		
					CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A25	No Apparent Adverse Event		
							F26	No Health Consequences or Impact	A25	No Apparent Adverse Event		
					CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A25	No Apparent Adverse Event		
							F26	No Health Consequences or Impact	A25	No Apparent Adverse Event		
					CT233	Surgical	E2403	No Clinical Signs, Symptoms or Conditions	A25	No Apparent Adverse Event		
							F26	No Health Consequences or Impact	A25	No Apparent Adverse Event		

					CT2406	Implantable	E2403	No Clinical Signs, Symptoms or Conditions	A25	No Apparent Adverse Event		
							F26	No Health Consequences or Impact	A25	No Apparent Adverse Event		
					CT2407	Active implantable	E2403	No Clinical Signs, Symptoms or Conditions	A25	No Apparent Adverse Event		
							F26	No Health Consequences or Impact	A25	No Apparent Adverse Event		
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A25	No Apparent Adverse Event		
							F26	No Health Consequences or Impact	A25	No Apparent Adverse Event		
					CT2727	Active implant devices/ancillaries	E2403	No Clinical Signs, Symptoms or Conditions	A25	No Apparent Adverse Event		
							F26	No Health Consequences or Impact	A25	No Apparent Adverse Event		
					CT2750	Implantable cardiac electrical stimulators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A25	No Apparent Adverse Event		
							F26	No Health Consequences or Impact	A25	No Apparent Adverse Event		
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A25	No Apparent Adverse Event		
							F26	No Health Consequences or Impact	A25	No Apparent Adverse Event		
					CT2985	Active	E2403	No Clinical Signs, Symptoms or Conditions	A25	No Apparent Adverse Event		
							F26	No Health Consequences or Impact	A25	No Apparent Adverse Event		
					CT3001	Implantable metal	E2403	No Clinical Signs, Symptoms or Conditions	A25	No Apparent Adverse Event		
							F26	No Health Consequences or Impact	A25	No Apparent Adverse Event		
					CT3002	Implantable synthetic polymer	E2403	No Clinical Signs, Symptoms or Conditions	A25	No Apparent Adverse Event		
							F26	No Health Consequences or Impact	A25	No Apparent Adverse Event		
					CT3006	Implantable inorganic material	E2403	No Clinical Signs, Symptoms or Conditions	A25	No Apparent Adverse Event		
							F26	No Health Consequences or Impact	A25	No Apparent Adverse Event		
					CT314	Electrically powered	E2403	No Clinical Signs, Symptoms or Conditions	A25	No Apparent Adverse Event		
							F26	No Health Consequences or Impact	A25	No Apparent Adverse Event		
					CT321	Long-term surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A25	No Apparent Adverse Event		
							F26	No Health Consequences or Impact	A25	No Apparent Adverse Event		
					CT449	Generators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A25	No Apparent Adverse Event		
							F26	No Health Consequences or Impact	A25	No Apparent Adverse Event		
					CT981	Single use	E2403	No Clinical Signs, Symptoms or Conditions	A25	No Apparent Adverse Event		
							F26	No Health Consequences or Impact	A25	No Apparent Adverse Event		
					CT983	Surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A25	No Apparent Adverse Event		
							F26	No Health Consequences or Impact	A25	No Apparent Adverse Event		
10618	2021	All other reportable incidents	36241	Pacemaker/defibrillator lead	CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A05	Mechanical Problem	21/01/2021	21/01/2021
									A070101	Failure to Capture	21/01/2021	21/01/2021
							F08	Hospitalization or Prolonged Hospitalization	A05	Mechanical Problem	21/01/2021	21/01/2021
									A070101	Failure to Capture	21/01/2021	21/01/2021
							F12	Serious Injury/ Illness/ Impairment	A05	Mechanical Problem	21/01/2021	21/01/2021
									A070101	Failure to Capture	21/01/2021	21/01/2021
							F1903	Device Explantation	A05	Mechanical Problem	21/01/2021	21/01/2021
									A070101	Failure to Capture	21/01/2021	21/01/2021
							F1905	Device Revision or Replacement	A05	Mechanical Problem	21/01/2021	21/01/2021
									A070101	Failure to Capture	21/01/2021	21/01/2021
10619	2021	All other reportable incidents	36241	Pacemaker/defibrillator lead	CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A05	Mechanical Problem	26/01/2021	26/01/2021
									A070102	High Capture Threshold	26/01/2021	26/01/2021
							F1905	Device Revision or Replacement	A05	Mechanical Problem	26/01/2021	26/01/2021
									A070102	High Capture Threshold	26/01/2021	26/01/2021
							F1908	Prolonged Surgery	A05	Mechanical Problem	26/01/2021	26/01/2021
									A070102	High Capture Threshold	26/01/2021	26/01/2021
							F26	No Health Consequences or Impact	A05	Mechanical Problem	26/01/2021	26/01/2021
									A070102	High Capture Threshold	26/01/2021	26/01/2021
10620	2021	All other reportable incidents	36241	Pacemaker/defibrillator lead	CT1684	Cardiac pacemakers and associated devices	E0604	Cardiac Perforation	A070101	Failure to Capture		
									A1301	Failure to Read Input Signal		
							E0619	Pericardial Effusion	A070101	Failure to Capture		
									A1301	Failure to Read Input Signal		
							E2114	Perforation	A070101	Failure to Capture		
									A1301	Failure to Read Input Signal		
							F08	Hospitalization or Prolonged Hospitalization	A070101	Failure to Capture		
									A1301	Failure to Read Input Signal		
							F12	Serious Injury/ Illness/ Impairment	A070101	Failure to Capture		
									A1301	Failure to Read Input Signal		
10621	2021	All other reportable incidents	36241	Pacemaker/defibrillator lead	CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A04	Material Integrity Problem	12/11/2020	12/11/2020
							F26	No Health Consequences or Impact	A04	Material Integrity Problem	12/11/2020	12/11/2020
10622	2021	All other reportable incidents	47265	Dual-chamber implantable pacemaker, rate-responsive	CT141	Electrodes/Leads and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem		
							F26	No Health Consequences or Impact	A0705	Battery Problem		
					CT145	Cardiovascular devices	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem		
							F26	No Health Consequences or Impact	A0705	Battery Problem		
					CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem		
							F26	No Health Consequences or Impact	A0705	Battery Problem		
					CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem		
							F26	No Health Consequences or Impact	A0705	Battery Problem		
					CT233	Surgical	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem		
							F26	No Health Consequences or Impact	A0705	Battery Problem		
					CT2406	Implantable	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem		
							F26	No Health Consequences or Impact	A0705	Battery Problem		
					CT2407	Active implantable	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem		
							F26	No Health Consequences or Impact	A0705	Battery Problem		
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem		
							F26	No Health Consequences or Impact	A0705	Battery Problem		
					CT2727	Active implant devices/ancillaries	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem		
							F26	No Health Consequences or Impact	A0705	Battery Problem		
					CT2750	Implantable cardiac electrical stimulators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem		
							F26	No Health Consequences or Impact	A0705	Battery Problem		
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem		
							F26	No Health Consequences or Impact	A0705	Battery Problem		
					CT2985	Active	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem		
							F26	No Health Consequences or Impact	A0705	Battery Problem		
					CT3001	Implantable metal	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem		
							F26	No Health Consequences or Impact	A0705	Battery Problem		
					CT3002	Implantable synthetic polymer	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem		
							F26	No Health Consequences or Impact	A0705	Battery Problem		

					CT3006	Implantable inorganic material	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem		
							F26	No Health Consequences or Impact	A0705	Battery Problem		
					CT314	Electrically powered	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem		
							F26	No Health Consequences or Impact	A0705	Battery Problem		
					CT321	Long-term surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem		
							F26	No Health Consequences or Impact	A0705	Battery Problem		
					CT449	Generators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem		
							F26	No Health Consequences or Impact	A0705	Battery Problem		
					CT981	Single use	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem		
							F26	No Health Consequences or Impact	A0705	Battery Problem		
					CT983	Surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem		
							F26	No Health Consequences or Impact	A0705	Battery Problem		
10623	2021	All other reportable incidents	47265	Dual-chamber implantable pacemaker, rate-responsive	CT141	Electrodes/Leads and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	04/02/2021	04/02/2021
							F19	Surgical Intervention	A070504	Premature Discharge of Battery	04/02/2021	04/02/2021
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	04/02/2021	04/02/2021
					CT145	Cardiovascular devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	04/02/2021	04/02/2021
							F19	Surgical Intervention	A070504	Premature Discharge of Battery	04/02/2021	04/02/2021
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	04/02/2021	04/02/2021
					CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	04/02/2021	04/02/2021
							F19	Surgical Intervention	A070504	Premature Discharge of Battery	04/02/2021	04/02/2021
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	04/02/2021	04/02/2021
					CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	04/02/2021	04/02/2021
							F19	Surgical Intervention	A070504	Premature Discharge of Battery	04/02/2021	04/02/2021
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	04/02/2021	04/02/2021
					CT233	Surgical	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	04/02/2021	04/02/2021
							F19	Surgical Intervention	A070504	Premature Discharge of Battery	04/02/2021	04/02/2021
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	04/02/2021	04/02/2021
					CT2406	Implantable	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	04/02/2021	04/02/2021
							F19	Surgical Intervention	A070504	Premature Discharge of Battery	04/02/2021	04/02/2021
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	04/02/2021	04/02/2021
					CT2407	Active implantable	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	04/02/2021	04/02/2021
							F19	Surgical Intervention	A070504	Premature Discharge of Battery	04/02/2021	04/02/2021
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	04/02/2021	04/02/2021
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	04/02/2021	04/02/2021
							F19	Surgical Intervention	A070504	Premature Discharge of Battery	04/02/2021	04/02/2021
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	04/02/2021	04/02/2021
					CT2727	Active implant devices/ancillaries	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	04/02/2021	04/02/2021
							F19	Surgical Intervention	A070504	Premature Discharge of Battery	04/02/2021	04/02/2021
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	04/02/2021	04/02/2021
					CT2750	Implantable cardiac electrical stimulators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	04/02/2021	04/02/2021
							F19	Surgical Intervention	A070504	Premature Discharge of Battery	04/02/2021	04/02/2021
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	04/02/2021	04/02/2021
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	04/02/2021	04/02/2021
							F19	Surgical Intervention	A070504	Premature Discharge of Battery	04/02/2021	04/02/2021
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	04/02/2021	04/02/2021
					CT2985	Active	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	04/02/2021	04/02/2021
							F19	Surgical Intervention	A070504	Premature Discharge of Battery	04/02/2021	04/02/2021
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	04/02/2021	04/02/2021
					CT3001	Implantable metal	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	04/02/2021	04/02/2021
							F19	Surgical Intervention	A070504	Premature Discharge of Battery	04/02/2021	04/02/2021
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	04/02/2021	04/02/2021
					CT3002	Implantable synthetic polymer	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	04/02/2021	04/02/2021
							F19	Surgical Intervention	A070504	Premature Discharge of Battery	04/02/2021	04/02/2021
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	04/02/2021	04/02/2021
					CT3006	Implantable inorganic material	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	04/02/2021	04/02/2021
							F19	Surgical Intervention	A070504	Premature Discharge of Battery	04/02/2021	04/02/2021
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	04/02/2021	04/02/2021
					CT314	Electrically powered	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	04/02/2021	04/02/2021
							F19	Surgical Intervention	A070504	Premature Discharge of Battery	04/02/2021	04/02/2021
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	04/02/2021	04/02/2021
					CT321	Long-term surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	04/02/2021	04/02/2021
							F19	Surgical Intervention	A070504	Premature Discharge of Battery	04/02/2021	04/02/2021
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	04/02/2021	04/02/2021
					CT449	Generators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	04/02/2021	04/02/2021
							F19	Surgical Intervention	A070504	Premature Discharge of Battery	04/02/2021	04/02/2021
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	04/02/2021	04/02/2021
					CT981	Single use	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	04/02/2021	04/02/2021
							F19	Surgical Intervention	A070504	Premature Discharge of Battery	04/02/2021	04/02/2021
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	04/02/2021	04/02/2021
					CT983	Surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	04/02/2021	04/02/2021
							F19	Surgical Intervention	A070504	Premature Discharge of Battery	04/02/2021	04/02/2021
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	04/02/2021	04/02/2021
10624	2021	All other reportable incidents	47263	Cardiac resynchronization therapy implantable pacemaker	CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A0908	Incorrect, Inadequate or Imprecise Result or Readings		
							F26	No Health Consequences or Impact	A0908	Incorrect, Inadequate or Imprecise Result or Readings		
					CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A0908	Incorrect, Inadequate or Imprecise Result or Readings		
							F26	No Health Consequences or Impact	A0908	Incorrect, Inadequate or Imprecise Result or Readings		
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A0908	Incorrect, Inadequate or Imprecise Result or Readings		
							F26	No Health Consequences or Impact	A0908	Incorrect, Inadequate or Imprecise Result or Readings		
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A0908	Incorrect, Inadequate or Imprecise Result or Readings		
							F26	No Health Consequences or Impact	A0908	Incorrect, Inadequate or Imprecise Result or Readings		
10625	2021	All other reportable incidents	35852	Single-chamber implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E2104	Electric Shock	A071302	Inappropriate/Inadequate Shock/Stimulation		
							F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A071302	Inappropriate/Inadequate Shock/Stimulation		

							F26	No Health Consequences or Impact	A071302	Inappropriate/Inadequate Shock/Stimulation		
10626	2021	All other reportable incidents	37265	Dual-chamber implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	23/02/2021	23/02/2021
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	23/02/2021	23/02/2021
							F19	Surgical Intervention	A070504	Premature Discharge of Battery	23/02/2021	23/02/2021
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	23/02/2021	23/02/2021
10627	2021	All other reportable incidents	35852	Single-chamber implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E2104	Electric Shock	A071302	Inappropriate/Inadequate Shock/Stimulation	26/04/2021	26/04/2021
							E2403	No Clinical Signs, Symptoms or Conditions	A071302	Inappropriate/Inadequate Shock/Stimulation	26/04/2021	26/04/2021
							F1905	Device Revision or Replacement	A071302	Inappropriate/Inadequate Shock/Stimulation	26/04/2021	26/04/2021
10628	2021	All other reportable incidents	35852	Single-chamber implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	28/01/2021	28/01/2021
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	28/01/2021	28/01/2021
							F19	Surgical Intervention	A070504	Premature Discharge of Battery	28/01/2021	28/01/2021
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	28/01/2021	28/01/2021
10629	2021	All other reportable incidents	37265	Dual-chamber implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A090805	Incorrect Measurement	11/02/2021	11/02/2021
							F1901	Additional Surgery	A090805	Incorrect Measurement	11/02/2021	11/02/2021
							F1905	Device Revision or Replacement	A090805	Incorrect Measurement	11/02/2021	11/02/2021
10630	2021	All other reportable incidents	35223	Endocardial/interventricular septal pacing lead	CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A0401	Break	05/02/2021	05/02/2021
									A0510	Retraction Problem	05/02/2021	05/02/2021
									A070101	Failure to Capture	05/02/2021	05/02/2021
									A072202	Low Impedance	05/02/2021	05/02/2021
							F1905	Device Revision or Replacement	A0401	Break	05/02/2021	05/02/2021
									A0510	Retraction Problem	05/02/2021	05/02/2021
									A070101	Failure to Capture	05/02/2021	05/02/2021
									A072202	Low Impedance	05/02/2021	05/02/2021
							F1908	Prolonged Surgery	A0401	Break	05/02/2021	05/02/2021
									A0510	Retraction Problem	05/02/2021	05/02/2021
									A070101	Failure to Capture	05/02/2021	05/02/2021
									A072202	Low Impedance	05/02/2021	05/02/2021
10631	2021	All other reportable incidents	36241	Pacemaker/defibrillator lead	CT1684	Cardiac pacemakers and associated devices	E1906	Unspecified Infection	A24	No Apparent Device Problem	02/02/2021	02/02/2021
							F08	Hospitalization or Prolonged Hospitalization	A24	No Apparent Device Problem	02/02/2021	02/02/2021
							F19	Surgical Intervention	A24	No Apparent Device Problem	02/02/2021	02/02/2021
							F1905	Device Revision or Replacement	A24	No Apparent Device Problem	02/02/2021	02/02/2021
10632	2021	All other reportable incidents	35852	Single-chamber implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A070910	Undersensing		
									A071201	Failure to Convert Rhythm		
							F05	Delay to Treatment/ Therapy	A070910	Undersensing		
									A071201	Failure to Convert Rhythm		
							F23	Unexpected Medical Intervention	A070910	Undersensing		
									A071201	Failure to Convert Rhythm		
10633	2021	All other reportable incidents	47263	Cardiac resynchronization therapy implantable pacemaker	CT1684	Cardiac pacemakers and associated devices	E1906	Unspecified Infection	A24	No Apparent Device Problem	02/02/2021	02/02/2021
							F08	Hospitalization or Prolonged Hospitalization	A24	No Apparent Device Problem	02/02/2021	02/02/2021
							F19	Surgical Intervention	A24	No Apparent Device Problem	02/02/2021	02/02/2021
					CT1685	Cardiac pacemakers	E1906	Unspecified Infection	A24	No Apparent Device Problem	02/02/2021	02/02/2021
							F08	Hospitalization or Prolonged Hospitalization	A24	No Apparent Device Problem	02/02/2021	02/02/2021
							F19	Surgical Intervention	A24	No Apparent Device Problem	02/02/2021	02/02/2021
					CT257	Implantable cardiac pacemakers	E1906	Unspecified Infection	A24	No Apparent Device Problem	02/02/2021	02/02/2021
							F08	Hospitalization or Prolonged Hospitalization	A24	No Apparent Device Problem	02/02/2021	02/02/2021
							F19	Surgical Intervention	A24	No Apparent Device Problem	02/02/2021	02/02/2021
					CT2866	Implantable cardiac electrical stimulators	E1906	Unspecified Infection	A24	No Apparent Device Problem	02/02/2021	02/02/2021
							F08	Hospitalization or Prolonged Hospitalization	A24	No Apparent Device Problem	02/02/2021	02/02/2021
							F19	Surgical Intervention	A24	No Apparent Device Problem	02/02/2021	02/02/2021
10634	2021	All other reportable incidents	47265	Dual-chamber implantable pacemaker, rate-responsive	CT141	Electrodes/Leads and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	14/03/2021	14/03/2021
							F19	Surgical Intervention	A070504	Premature Discharge of Battery	14/03/2021	14/03/2021
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	14/03/2021	14/03/2021
					CT145	Cardiovascular devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	14/03/2021	14/03/2021
							F19	Surgical Intervention	A070504	Premature Discharge of Battery	14/03/2021	14/03/2021
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	14/03/2021	14/03/2021
					CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	14/03/2021	14/03/2021
							F19	Surgical Intervention	A070504	Premature Discharge of Battery	14/03/2021	14/03/2021
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	14/03/2021	14/03/2021
					CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	14/03/2021	14/03/2021
							F19	Surgical Intervention	A070504	Premature Discharge of Battery	14/03/2021	14/03/2021
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	14/03/2021	14/03/2021
					CT233	Surgical	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	14/03/2021	14/03/2021
							F19	Surgical Intervention	A070504	Premature Discharge of Battery	14/03/2021	14/03/2021
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	14/03/2021	14/03/2021
					CT2406	Implantable	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	14/03/2021	14/03/2021
							F19	Surgical Intervention	A070504	Premature Discharge of Battery	14/03/2021	14/03/2021
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	14/03/2021	14/03/2021
					CT2407	Active implantable	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	14/03/2021	14/03/2021
							F19	Surgical Intervention	A070504	Premature Discharge of Battery	14/03/2021	14/03/2021
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	14/03/2021	14/03/2021
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	14/03/2021	14/03/2021
							F19	Surgical Intervention	A070504	Premature Discharge of Battery	14/03/2021	14/03/2021
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	14/03/2021	14/03/2021
					CT2727	Active implant devices/ancillaries	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	14/03/2021	14/03/2021
							F19	Surgical Intervention	A070504	Premature Discharge of Battery	14/03/2021	14/03/2021
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	14/03/2021	14/03/2021
					CT2750	Implantable cardiac electrical stimulators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	14/03/2021	14/03/2021
							F19	Surgical Intervention	A070504	Premature Discharge of Battery	14/03/2021	14/03/2021
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	14/03/2021	14/03/2021
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	14/03/2021	14/03/2021
							F19	Surgical Intervention	A070504	Premature Discharge of Battery	14/03/2021	14/03/2021
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	14/03/2021	14/03/2021
					CT2985	Active	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	14/03/2021	14/03/2021
							F19	Surgical Intervention	A070504	Premature Discharge of Battery	14/03/2021	14/03/2021
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	14/03/2021	14/03/2021
					CT3001	Implantable metal	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	14/03/2021	14/03/2021
							F19	Surgical Intervention	A070504	Premature Discharge of Battery	14/03/2021	14/03/2021
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	14/03/2021	14/03/2021

					CT3002	Implantable synthetic polymer	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	14/03/2021	14/03/2021
							F19	Surgical Intervention	A070504	Premature Discharge of Battery	14/03/2021	14/03/2021
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	14/03/2021	14/03/2021
					CT3006	Implantable inorganic material	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	14/03/2021	14/03/2021
							F19	Surgical Intervention	A070504	Premature Discharge of Battery	14/03/2021	14/03/2021
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	14/03/2021	14/03/2021
					CT314	Electrically powered	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	14/03/2021	14/03/2021
							F19	Surgical Intervention	A070504	Premature Discharge of Battery	14/03/2021	14/03/2021
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	14/03/2021	14/03/2021
					CT321	Long-term surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	14/03/2021	14/03/2021
							F19	Surgical Intervention	A070504	Premature Discharge of Battery	14/03/2021	14/03/2021
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	14/03/2021	14/03/2021
					CT449	Generators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	14/03/2021	14/03/2021
							F19	Surgical Intervention	A070504	Premature Discharge of Battery	14/03/2021	14/03/2021
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	14/03/2021	14/03/2021
					CT981	Single use	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	14/03/2021	14/03/2021
							F19	Surgical Intervention	A070504	Premature Discharge of Battery	14/03/2021	14/03/2021
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	14/03/2021	14/03/2021
					CT983	Surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	14/03/2021	14/03/2021
							F19	Surgical Intervention	A070504	Premature Discharge of Battery	14/03/2021	14/03/2021
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	14/03/2021	14/03/2021
10635	2021	All other reportable incidents	47265	Dual-chamber implantable pacemaker, rate-responsive	CT141	Electrodes/Leads and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A0908	Incorrect, Inadequate or Imprecise Result or Readings		
							F08	Hospitalization or Prolonged Hospitalization	A0908	Incorrect, Inadequate or Imprecise Result or Readings		
							F1905	Device Revision or Replacement	A0908	Incorrect, Inadequate or Imprecise Result or Readings		
					CT145	Cardiovascular devices	E2403	No Clinical Signs, Symptoms or Conditions	A0908	Incorrect, Inadequate or Imprecise Result or Readings		
							F08	Hospitalization or Prolonged Hospitalization	A0908	Incorrect, Inadequate or Imprecise Result or Readings		
							F1905	Device Revision or Replacement	A0908	Incorrect, Inadequate or Imprecise Result or Readings		
					CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A0908	Incorrect, Inadequate or Imprecise Result or Readings		
							F08	Hospitalization or Prolonged Hospitalization	A0908	Incorrect, Inadequate or Imprecise Result or Readings		
							F1905	Device Revision or Replacement	A0908	Incorrect, Inadequate or Imprecise Result or Readings		
					CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A0908	Incorrect, Inadequate or Imprecise Result or Readings		
							F08	Hospitalization or Prolonged Hospitalization	A0908	Incorrect, Inadequate or Imprecise Result or Readings		
							F1905	Device Revision or Replacement	A0908	Incorrect, Inadequate or Imprecise Result or Readings		
					CT233	Surgical	E2403	No Clinical Signs, Symptoms or Conditions	A0908	Incorrect, Inadequate or Imprecise Result or Readings		
							F08	Hospitalization or Prolonged Hospitalization	A0908	Incorrect, Inadequate or Imprecise Result or Readings		
							F1905	Device Revision or Replacement	A0908	Incorrect, Inadequate or Imprecise Result or Readings		
					CT2406	Implantable	E2403	No Clinical Signs, Symptoms or Conditions	A0908	Incorrect, Inadequate or Imprecise Result or Readings		
							F08	Hospitalization or Prolonged Hospitalization	A0908	Incorrect, Inadequate or Imprecise Result or Readings		
							F1905	Device Revision or Replacement	A0908	Incorrect, Inadequate or Imprecise Result or Readings		
					CT2407	Active implantable	E2403	No Clinical Signs, Symptoms or Conditions	A0908	Incorrect, Inadequate or Imprecise Result or Readings		
							F08	Hospitalization or Prolonged Hospitalization	A0908	Incorrect, Inadequate or Imprecise Result or Readings		
							F1905	Device Revision or Replacement	A0908	Incorrect, Inadequate or Imprecise Result or Readings		
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A0908	Incorrect, Inadequate or Imprecise Result or Readings		
							F08	Hospitalization or Prolonged Hospitalization	A0908	Incorrect, Inadequate or Imprecise Result or Readings		
							F1905	Device Revision or Replacement	A0908	Incorrect, Inadequate or Imprecise Result or Readings		
					CT2727	Active implant devices/ancillaries	E2403	No Clinical Signs, Symptoms or Conditions	A0908	Incorrect, Inadequate or Imprecise Result or Readings		
							F08	Hospitalization or Prolonged Hospitalization	A0908	Incorrect, Inadequate or Imprecise Result or Readings		
							F1905	Device Revision or Replacement	A0908	Incorrect, Inadequate or Imprecise Result or Readings		
					CT2750	Implantable cardiac electrical stimulators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A0908	Incorrect, Inadequate or Imprecise Result or Readings		
							F08	Hospitalization or Prolonged Hospitalization	A0908	Incorrect, Inadequate or Imprecise Result or Readings		
							F1905	Device Revision or Replacement	A0908	Incorrect, Inadequate or Imprecise Result or Readings		
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A0908	Incorrect, Inadequate or Imprecise Result or Readings		
							F08	Hospitalization or Prolonged Hospitalization	A0908	Incorrect, Inadequate or Imprecise Result or Readings		
							F1905	Device Revision or Replacement	A0908	Incorrect, Inadequate or Imprecise Result or Readings		
					CT2985	Active	E2403	No Clinical Signs, Symptoms or Conditions	A0908	Incorrect, Inadequate or Imprecise Result or Readings		
							F08	Hospitalization or Prolonged Hospitalization	A0908	Incorrect, Inadequate or Imprecise Result or Readings		
							F1905	Device Revision or Replacement	A0908	Incorrect, Inadequate or Imprecise Result or Readings		
					CT3001	Implantable metal	E2403	No Clinical Signs, Symptoms or Conditions	A0908	Incorrect, Inadequate or Imprecise Result or Readings		
							F08	Hospitalization or Prolonged Hospitalization	A0908	Incorrect, Inadequate or Imprecise Result or Readings		

							F1905	Device Revision or Replacement	A0908	Incorrect, Inadequate or Imprecise Result or Readings			
					CT3002	Implantable synthetic polymer	E2403	No Clinical Signs, Symptoms or Conditions	A0908	Incorrect, Inadequate or Imprecise Result or Readings			
							F08	Hospitalization or Prolonged Hospitalization	A0908	Incorrect, Inadequate or Imprecise Result or Readings			
							F1905	Device Revision or Replacement	A0908	Incorrect, Inadequate or Imprecise Result or Readings			
					CT3006	Implantable inorganic material	E2403	No Clinical Signs, Symptoms or Conditions	A0908	Incorrect, Inadequate or Imprecise Result or Readings			
							F08	Hospitalization or Prolonged Hospitalization	A0908	Incorrect, Inadequate or Imprecise Result or Readings			
							F1905	Device Revision or Replacement	A0908	Incorrect, Inadequate or Imprecise Result or Readings			
					CT314	Electrically powered	E2403	No Clinical Signs, Symptoms or Conditions	A0908	Incorrect, Inadequate or Imprecise Result or Readings			
							F08	Hospitalization or Prolonged Hospitalization	A0908	Incorrect, Inadequate or Imprecise Result or Readings			
							F1905	Device Revision or Replacement	A0908	Incorrect, Inadequate or Imprecise Result or Readings			
					CT321	Long-term surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A0908	Incorrect, Inadequate or Imprecise Result or Readings			
							F08	Hospitalization or Prolonged Hospitalization	A0908	Incorrect, Inadequate or Imprecise Result or Readings			
							F1905	Device Revision or Replacement	A0908	Incorrect, Inadequate or Imprecise Result or Readings			
					CT449	Generators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A0908	Incorrect, Inadequate or Imprecise Result or Readings			
							F08	Hospitalization or Prolonged Hospitalization	A0908	Incorrect, Inadequate or Imprecise Result or Readings			
							F1905	Device Revision or Replacement	A0908	Incorrect, Inadequate or Imprecise Result or Readings			
					CT981	Single use	E2403	No Clinical Signs, Symptoms or Conditions	A0908	Incorrect, Inadequate or Imprecise Result or Readings			
							F08	Hospitalization or Prolonged Hospitalization	A0908	Incorrect, Inadequate or Imprecise Result or Readings			
							F1905	Device Revision or Replacement	A0908	Incorrect, Inadequate or Imprecise Result or Readings			
					CT983	Surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A0908	Incorrect, Inadequate or Imprecise Result or Readings			
							F08	Hospitalization or Prolonged Hospitalization	A0908	Incorrect, Inadequate or Imprecise Result or Readings			
							F1905	Device Revision or Replacement	A0908	Incorrect, Inadequate or Imprecise Result or Readings			
10636	2021	All other reportable incidents	47265	Dual-chamber implantable pacemaker, rate-responsive	CT141	Electrodes/Leads and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem	28/04/2021	28/04/2021
									A070504	Premature Discharge of Battery	28/04/2021	28/04/2021
							F08	Hospitalization or Prolonged Hospitalization	A0705	Battery Problem	28/04/2021	28/04/2021
									A070504	Premature Discharge of Battery	28/04/2021	28/04/2021
							F1905	Device Revision or Replacement	A0705	Battery Problem	28/04/2021	28/04/2021
									A070504	Premature Discharge of Battery	28/04/2021	28/04/2021
					CT145	Cardiovascular devices	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem	28/04/2021	28/04/2021
									A070504	Premature Discharge of Battery	28/04/2021	28/04/2021
							F08	Hospitalization or Prolonged Hospitalization	A0705	Battery Problem	28/04/2021	28/04/2021
									A070504	Premature Discharge of Battery	28/04/2021	28/04/2021
							F1905	Device Revision or Replacement	A0705	Battery Problem	28/04/2021	28/04/2021
									A070504	Premature Discharge of Battery	28/04/2021	28/04/2021
					CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem	28/04/2021	28/04/2021
									A070504	Premature Discharge of Battery	28/04/2021	28/04/2021
							F08	Hospitalization or Prolonged Hospitalization	A0705	Battery Problem	28/04/2021	28/04/2021
									A070504	Premature Discharge of Battery	28/04/2021	28/04/2021
							F1905	Device Revision or Replacement	A0705	Battery Problem	28/04/2021	28/04/2021
									A070504	Premature Discharge of Battery	28/04/2021	28/04/2021
					CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem	28/04/2021	28/04/2021
									A070504	Premature Discharge of Battery	28/04/2021	28/04/2021
							F08	Hospitalization or Prolonged Hospitalization	A0705	Battery Problem	28/04/2021	28/04/2021
									A070504	Premature Discharge of Battery	28/04/2021	28/04/2021
							F1905	Device Revision or Replacement	A0705	Battery Problem	28/04/2021	28/04/2021
									A070504	Premature Discharge of Battery	28/04/2021	28/04/2021
					CT233	Surgical	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem	28/04/2021	28/04/2021
									A070504	Premature Discharge of Battery	28/04/2021	28/04/2021
							F08	Hospitalization or Prolonged Hospitalization	A0705	Battery Problem	28/04/2021	28/04/2021
									A070504	Premature Discharge of Battery	28/04/2021	28/04/2021
							F1905	Device Revision or Replacement	A0705	Battery Problem	28/04/2021	28/04/2021
									A070504	Premature Discharge of Battery	28/04/2021	28/04/2021
					CT2406	Implantable	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem	28/04/2021	28/04/2021
									A070504	Premature Discharge of Battery	28/04/2021	28/04/2021
							F08	Hospitalization or Prolonged Hospitalization	A0705	Battery Problem	28/04/2021	28/04/2021
									A070504	Premature Discharge of Battery	28/04/2021	28/04/2021
							F1905	Device Revision or Replacement	A0705	Battery Problem	28/04/2021	28/04/2021
									A070504	Premature Discharge of Battery	28/04/2021	28/04/2021
					CT2407	Active implantable	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem	28/04/2021	28/04/2021
									A070504	Premature Discharge of Battery	28/04/2021	28/04/2021
							F08	Hospitalization or Prolonged Hospitalization	A0705	Battery Problem	28/04/2021	28/04/2021
									A070504	Premature Discharge of Battery	28/04/2021	28/04/2021
							F1905	Device Revision or Replacement	A0705	Battery Problem	28/04/2021	28/04/2021
									A070504	Premature Discharge of Battery	28/04/2021	28/04/2021
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem	28/04/2021	28/04/2021
									A070504	Premature Discharge of Battery	28/04/2021	28/04/2021
							F08	Hospitalization or Prolonged Hospitalization	A0705	Battery Problem	28/04/2021	28/04/2021
									A070504	Premature Discharge of Battery	28/04/2021	28/04/2021
							F1905	Device Revision or Replacement	A0705	Battery Problem	28/04/2021	28/04/2021
									A070504	Premature Discharge of Battery	28/04/2021	28/04/2021
					CT2727	Active implant devices/ancillaries	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem	28/04/2021	28/04/2021

								A070504	Premature Discharge of Battery	28/04/2021	28/04/2021
						F08	Hospitalization or Prolonged Hospitalization	A0705	Battery Problem	28/04/2021	28/04/2021
								A070504	Premature Discharge of Battery	28/04/2021	28/04/2021
						F1905	Device Revision or Replacement	A0705	Battery Problem	28/04/2021	28/04/2021
								A070504	Premature Discharge of Battery	28/04/2021	28/04/2021
				CT2750	Implantable cardiac electrical stimulators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem	28/04/2021	28/04/2021
								A070504	Premature Discharge of Battery	28/04/2021	28/04/2021
						F08	Hospitalization or Prolonged Hospitalization	A0705	Battery Problem	28/04/2021	28/04/2021
								A070504	Premature Discharge of Battery	28/04/2021	28/04/2021
						F1905	Device Revision or Replacement	A0705	Battery Problem	28/04/2021	28/04/2021
								A070504	Premature Discharge of Battery	28/04/2021	28/04/2021
				CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem	28/04/2021	28/04/2021
								A070504	Premature Discharge of Battery	28/04/2021	28/04/2021
						F08	Hospitalization or Prolonged Hospitalization	A0705	Battery Problem	28/04/2021	28/04/2021
								A070504	Premature Discharge of Battery	28/04/2021	28/04/2021
						F1905	Device Revision or Replacement	A0705	Battery Problem	28/04/2021	28/04/2021
								A070504	Premature Discharge of Battery	28/04/2021	28/04/2021
				CT2985	Active	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem	28/04/2021	28/04/2021
								A070504	Premature Discharge of Battery	28/04/2021	28/04/2021
						F08	Hospitalization or Prolonged Hospitalization	A0705	Battery Problem	28/04/2021	28/04/2021
								A070504	Premature Discharge of Battery	28/04/2021	28/04/2021
						F1905	Device Revision or Replacement	A0705	Battery Problem	28/04/2021	28/04/2021
								A070504	Premature Discharge of Battery	28/04/2021	28/04/2021
				CT3001	Implantable metal	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem	28/04/2021	28/04/2021
								A070504	Premature Discharge of Battery	28/04/2021	28/04/2021
						F08	Hospitalization or Prolonged Hospitalization	A0705	Battery Problem	28/04/2021	28/04/2021
								A070504	Premature Discharge of Battery	28/04/2021	28/04/2021
						F1905	Device Revision or Replacement	A0705	Battery Problem	28/04/2021	28/04/2021
								A070504	Premature Discharge of Battery	28/04/2021	28/04/2021
				CT3002	Implantable synthetic polymer	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem	28/04/2021	28/04/2021
								A070504	Premature Discharge of Battery	28/04/2021	28/04/2021
						F08	Hospitalization or Prolonged Hospitalization	A0705	Battery Problem	28/04/2021	28/04/2021
								A070504	Premature Discharge of Battery	28/04/2021	28/04/2021
						F1905	Device Revision or Replacement	A0705	Battery Problem	28/04/2021	28/04/2021
								A070504	Premature Discharge of Battery	28/04/2021	28/04/2021
				CT3006	Implantable inorganic material	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem	28/04/2021	28/04/2021
								A070504	Premature Discharge of Battery	28/04/2021	28/04/2021
						F08	Hospitalization or Prolonged Hospitalization	A0705	Battery Problem	28/04/2021	28/04/2021
								A070504	Premature Discharge of Battery	28/04/2021	28/04/2021
						F1905	Device Revision or Replacement	A0705	Battery Problem	28/04/2021	28/04/2021
								A070504	Premature Discharge of Battery	28/04/2021	28/04/2021
				CT314	Electrically powered	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem	28/04/2021	28/04/2021
								A070504	Premature Discharge of Battery	28/04/2021	28/04/2021
						F08	Hospitalization or Prolonged Hospitalization	A0705	Battery Problem	28/04/2021	28/04/2021
								A070504	Premature Discharge of Battery	28/04/2021	28/04/2021
						F1905	Device Revision or Replacement	A0705	Battery Problem	28/04/2021	28/04/2021
								A070504	Premature Discharge of Battery	28/04/2021	28/04/2021
				CT321	Long-term surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem	28/04/2021	28/04/2021
								A070504	Premature Discharge of Battery	28/04/2021	28/04/2021
						F08	Hospitalization or Prolonged Hospitalization	A0705	Battery Problem	28/04/2021	28/04/2021
								A070504	Premature Discharge of Battery	28/04/2021	28/04/2021
						F1905	Device Revision or Replacement	A0705	Battery Problem	28/04/2021	28/04/2021
								A070504	Premature Discharge of Battery	28/04/2021	28/04/2021
				CT449	Generators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem	28/04/2021	28/04/2021
								A070504	Premature Discharge of Battery	28/04/2021	28/04/2021
						F08	Hospitalization or Prolonged Hospitalization	A0705	Battery Problem	28/04/2021	28/04/2021
								A070504	Premature Discharge of Battery	28/04/2021	28/04/2021
						F1905	Device Revision or Replacement	A0705	Battery Problem	28/04/2021	28/04/2021
								A070504	Premature Discharge of Battery	28/04/2021	28/04/2021
				CT981	Single use	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem	28/04/2021	28/04/2021
								A070504	Premature Discharge of Battery	28/04/2021	28/04/2021
						F08	Hospitalization or Prolonged Hospitalization	A0705	Battery Problem	28/04/2021	28/04/2021
								A070504	Premature Discharge of Battery	28/04/2021	28/04/2021
						F1905	Device Revision or Replacement	A0705	Battery Problem	28/04/2021	28/04/2021
								A070504	Premature Discharge of Battery	28/04/2021	28/04/2021
				CT983	Surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem	28/04/2021	28/04/2021
								A070504	Premature Discharge of Battery	28/04/2021	28/04/2021
						F08	Hospitalization or Prolonged Hospitalization	A0705	Battery Problem	28/04/2021	28/04/2021
								A070504	Premature Discharge of Battery	28/04/2021	28/04/2021
						F1905	Device Revision or Replacement	A0705	Battery Problem	28/04/2021	28/04/2021
								A070504	Premature Discharge of Battery	28/04/2021	28/04/2021
10637	2021	All other reportable incidents	47265	Dual-chamber implantable pacemaker, rate-responsive	CT141	Electrodes/Leads and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem	28/03/2021	28/03/2021
									A070504	Premature Discharge of Battery	28/03/2021	28/03/2021
						F1903	Device Explantation	A0705	Battery Problem	28/03/2021	28/03/2021
								A070504	Premature Discharge of Battery	28/03/2021	28/03/2021
						F1905	Device Revision or Replacement	A0705	Battery Problem	28/03/2021	28/03/2021
								A070504	Premature Discharge of Battery	28/03/2021	28/03/2021
				CT145	Cardiovascular devices	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem	28/03/2021	28/03/2021
								A070504	Premature Discharge of Battery	28/03/2021	28/03/2021
						F1903	Device Explantation	A0705	Battery Problem	28/03/2021	28/03/2021
								A070504	Premature Discharge of Battery	28/03/2021	28/03/2021
						F1905	Device Revision or Replacement	A0705	Battery Problem	28/03/2021	28/03/2021
								A070504	Premature Discharge of Battery	28/03/2021	28/03/2021
				CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem	28/03/2021	28/03/2021
								A070504	Premature Discharge of Battery	28/03/2021	28/03/2021
						F1903	Device Explantation	A0705	Battery Problem	28/03/2021	28/03/2021
								A070504	Premature Discharge of Battery	28/03/2021	28/03/2021
						F1905	Device Revision or Replacement	A0705	Battery Problem	28/03/2021	28/03/2021

								A070504	Premature Discharge of Battery	28/03/2021	28/03/2021
					CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem	28/03/2021	28/03/2021
									A070504	Premature Discharge of Battery	28/03/2021	28/03/2021
							F1903	Device Explantation	A0705	Battery Problem	28/03/2021	28/03/2021
									A070504	Premature Discharge of Battery	28/03/2021	28/03/2021
							F1905	Device Revision or Replacement	A0705	Battery Problem	28/03/2021	28/03/2021
									A070504	Premature Discharge of Battery	28/03/2021	28/03/2021
					CT233	Surgical	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem	28/03/2021	28/03/2021
									A070504	Premature Discharge of Battery	28/03/2021	28/03/2021
							F1903	Device Explantation	A0705	Battery Problem	28/03/2021	28/03/2021
									A070504	Premature Discharge of Battery	28/03/2021	28/03/2021
							F1905	Device Revision or Replacement	A0705	Battery Problem	28/03/2021	28/03/2021
									A070504	Premature Discharge of Battery	28/03/2021	28/03/2021
					CT2406	Implantable	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem	28/03/2021	28/03/2021
									A070504	Premature Discharge of Battery	28/03/2021	28/03/2021
							F1903	Device Explantation	A0705	Battery Problem	28/03/2021	28/03/2021
									A070504	Premature Discharge of Battery	28/03/2021	28/03/2021
							F1905	Device Revision or Replacement	A0705	Battery Problem	28/03/2021	28/03/2021
									A070504	Premature Discharge of Battery	28/03/2021	28/03/2021
					CT2407	Active implantable	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem	28/03/2021	28/03/2021
									A070504	Premature Discharge of Battery	28/03/2021	28/03/2021
							F1903	Device Explantation	A0705	Battery Problem	28/03/2021	28/03/2021
									A070504	Premature Discharge of Battery	28/03/2021	28/03/2021
							F1905	Device Revision or Replacement	A0705	Battery Problem	28/03/2021	28/03/2021
									A070504	Premature Discharge of Battery	28/03/2021	28/03/2021
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem	28/03/2021	28/03/2021
									A070504	Premature Discharge of Battery	28/03/2021	28/03/2021
							F1903	Device Explantation	A0705	Battery Problem	28/03/2021	28/03/2021
									A070504	Premature Discharge of Battery	28/03/2021	28/03/2021
							F1905	Device Revision or Replacement	A0705	Battery Problem	28/03/2021	28/03/2021
									A070504	Premature Discharge of Battery	28/03/2021	28/03/2021
					CT2727	Active implant devices/ancillaries	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem	28/03/2021	28/03/2021
									A070504	Premature Discharge of Battery	28/03/2021	28/03/2021
							F1903	Device Explantation	A0705	Battery Problem	28/03/2021	28/03/2021
									A070504	Premature Discharge of Battery	28/03/2021	28/03/2021
							F1905	Device Revision or Replacement	A0705	Battery Problem	28/03/2021	28/03/2021
									A070504	Premature Discharge of Battery	28/03/2021	28/03/2021
					CT2750	Implantable cardiac electrical stimulators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem	28/03/2021	28/03/2021
									A070504	Premature Discharge of Battery	28/03/2021	28/03/2021
							F1903	Device Explantation	A0705	Battery Problem	28/03/2021	28/03/2021
									A070504	Premature Discharge of Battery	28/03/2021	28/03/2021
							F1905	Device Revision or Replacement	A0705	Battery Problem	28/03/2021	28/03/2021
									A070504	Premature Discharge of Battery	28/03/2021	28/03/2021
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem	28/03/2021	28/03/2021
									A070504	Premature Discharge of Battery	28/03/2021	28/03/2021
							F1903	Device Explantation	A0705	Battery Problem	28/03/2021	28/03/2021
									A070504	Premature Discharge of Battery	28/03/2021	28/03/2021
							F1905	Device Revision or Replacement	A0705	Battery Problem	28/03/2021	28/03/2021
									A070504	Premature Discharge of Battery	28/03/2021	28/03/2021
					CT2985	Active	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem	28/03/2021	28/03/2021
									A070504	Premature Discharge of Battery	28/03/2021	28/03/2021
							F1903	Device Explantation	A0705	Battery Problem	28/03/2021	28/03/2021
									A070504	Premature Discharge of Battery	28/03/2021	28/03/2021
							F1905	Device Revision or Replacement	A0705	Battery Problem	28/03/2021	28/03/2021
									A070504	Premature Discharge of Battery	28/03/2021	28/03/2021
					CT3001	Implantable metal	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem	28/03/2021	28/03/2021
									A070504	Premature Discharge of Battery	28/03/2021	28/03/2021
							F1903	Device Explantation	A0705	Battery Problem	28/03/2021	28/03/2021
									A070504	Premature Discharge of Battery	28/03/2021	28/03/2021
							F1905	Device Revision or Replacement	A0705	Battery Problem	28/03/2021	28/03/2021
									A070504	Premature Discharge of Battery	28/03/2021	28/03/2021
					CT3002	Implantable synthetic polymer	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem	28/03/2021	28/03/2021
									A070504	Premature Discharge of Battery	28/03/2021	28/03/2021
							F1903	Device Explantation	A0705	Battery Problem	28/03/2021	28/03/2021
									A070504	Premature Discharge of Battery	28/03/2021	28/03/2021
							F1905	Device Revision or Replacement	A0705	Battery Problem	28/03/2021	28/03/2021
									A070504	Premature Discharge of Battery	28/03/2021	28/03/2021
					CT3006	Implantable inorganic material	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem	28/03/2021	28/03/2021
									A070504	Premature Discharge of Battery	28/03/2021	28/03/2021
							F1903	Device Explantation	A0705	Battery Problem	28/03/2021	28/03/2021
									A070504	Premature Discharge of Battery	28/03/2021	28/03/2021
							F1905	Device Revision or Replacement	A0705	Battery Problem	28/03/2021	28/03/2021
									A070504	Premature Discharge of Battery	28/03/2021	28/03/2021
					CT314	Electrically powered	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem	28/03/2021	28/03/2021
									A070504	Premature Discharge of Battery	28/03/2021	28/03/2021
							F1903	Device Explantation	A0705	Battery Problem	28/03/2021	28/03/2021
									A070504	Premature Discharge of Battery	28/03/2021	28/03/2021
							F1905	Device Revision or Replacement	A0705	Battery Problem	28/03/2021	28/03/2021
									A070504	Premature Discharge of Battery	28/03/2021	28/03/2021
					CT321	Long-term surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem	28/03/2021	28/03/2021
									A070504	Premature Discharge of Battery	28/03/2021	28/03/2021
							F1903	Device Explantation	A0705	Battery Problem	28/03/2021	28/03/2021
									A070504	Premature Discharge of Battery	28/03/2021	28/03/2021
							F1905	Device Revision or Replacement	A0705	Battery Problem	28/03/2021	28/03/2021
									A070504	Premature Discharge of Battery	28/03/2021	28/03/2021
					CT449	Generators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem	28/03/2021	28/03/2021
									A070504	Premature Discharge of Battery	28/03/2021	28/03/2021
							F1903	Device Explantation	A0705	Battery Problem	28/03/2021	28/03/2021

								A070504	Premature Discharge of Battery	28/03/2021	28/03/2021
						F1905	Device Revision or Replacement	A0705	Battery Problem	28/03/2021	28/03/2021
								A070504	Premature Discharge of Battery	28/03/2021	28/03/2021
				CT981	Single use	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem	28/03/2021	28/03/2021
								A070504	Premature Discharge of Battery	28/03/2021	28/03/2021
						F1903	Device Explantation	A0705	Battery Problem	28/03/2021	28/03/2021
								A070504	Premature Discharge of Battery	28/03/2021	28/03/2021
						F1905	Device Revision or Replacement	A0705	Battery Problem	28/03/2021	28/03/2021
								A070504	Premature Discharge of Battery	28/03/2021	28/03/2021
				CT983	Surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem	28/03/2021	28/03/2021
								A070504	Premature Discharge of Battery	28/03/2021	28/03/2021
						F1903	Device Explantation	A0705	Battery Problem	28/03/2021	28/03/2021
								A070504	Premature Discharge of Battery	28/03/2021	28/03/2021
						F1905	Device Revision or Replacement	A0705	Battery Problem	28/03/2021	28/03/2021
								A070504	Premature Discharge of Battery	28/03/2021	28/03/2021
10638	2021	All other reportable incidents	35852	Single-chamber implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E2104	Electric Shock	A070909	Oversensing		
									A071302	Inappropriate/Inadequate Shock/Stimulation		
							F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A070909	Oversensing		
									A071302	Inappropriate/Inadequate Shock/Stimulation		
							F26	No Health Consequences or Impact	A070909	Oversensing		
									A071302	Inappropriate/Inadequate Shock/Stimulation		
10639	2021	All other reportable incidents	47263	Cardiac resynchronization therapy implantable pacemaker	CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A0908	Incorrect, Inadequate or Imprecise Result or Readings	29/01/2021	29/01/2021
							F08	Hospitalization or Prolonged Hospitalization	A0908	Incorrect, Inadequate or Imprecise Result or Readings	29/01/2021	29/01/2021
							F12	Serious Injury/ Illness/ Impairment	A0908	Incorrect, Inadequate or Imprecise Result or Readings	29/01/2021	29/01/2021
							F19	Surgical Intervention	A0908	Incorrect, Inadequate or Imprecise Result or Readings	29/01/2021	29/01/2021
							F1905	Device Revision or Replacement	A0908	Incorrect, Inadequate or Imprecise Result or Readings	29/01/2021	29/01/2021
				CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A0908	Incorrect, Inadequate or Imprecise Result or Readings	29/01/2021	29/01/2021
							F08	Hospitalization or Prolonged Hospitalization	A0908	Incorrect, Inadequate or Imprecise Result or Readings	29/01/2021	29/01/2021
							F12	Serious Injury/ Illness/ Impairment	A0908	Incorrect, Inadequate or Imprecise Result or Readings	29/01/2021	29/01/2021
							F19	Surgical Intervention	A0908	Incorrect, Inadequate or Imprecise Result or Readings	29/01/2021	29/01/2021
							F1905	Device Revision or Replacement	A0908	Incorrect, Inadequate or Imprecise Result or Readings	29/01/2021	29/01/2021
				CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A0908	Incorrect, Inadequate or Imprecise Result or Readings	29/01/2021	29/01/2021
							F08	Hospitalization or Prolonged Hospitalization	A0908	Incorrect, Inadequate or Imprecise Result or Readings	29/01/2021	29/01/2021
							F12	Serious Injury/ Illness/ Impairment	A0908	Incorrect, Inadequate or Imprecise Result or Readings	29/01/2021	29/01/2021
							F19	Surgical Intervention	A0908	Incorrect, Inadequate or Imprecise Result or Readings	29/01/2021	29/01/2021
							F1905	Device Revision or Replacement	A0908	Incorrect, Inadequate or Imprecise Result or Readings	29/01/2021	29/01/2021
				CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A0908	Incorrect, Inadequate or Imprecise Result or Readings	29/01/2021	29/01/2021
							F08	Hospitalization or Prolonged Hospitalization	A0908	Incorrect, Inadequate or Imprecise Result or Readings	29/01/2021	29/01/2021
							F12	Serious Injury/ Illness/ Impairment	A0908	Incorrect, Inadequate or Imprecise Result or Readings	29/01/2021	29/01/2021
							F19	Surgical Intervention	A0908	Incorrect, Inadequate or Imprecise Result or Readings	29/01/2021	29/01/2021
							F1905	Device Revision or Replacement	A0908	Incorrect, Inadequate or Imprecise Result or Readings	29/01/2021	29/01/2021
10640	2021	All other reportable incidents	37265	Dual-chamber implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A070909	Oversensing		
							F26	No Health Consequences or Impact	A070909	Oversensing		
10641	2021	All other reportable incidents	47265	Dual-chamber implantable pacemaker, rate-responsive	CT141	Electrodes/Leads and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	04/05/2021	04/05/2021
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	04/05/2021	04/05/2021
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	04/05/2021	04/05/2021
				CT145	Cardiovascular devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	04/05/2021	04/05/2021
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	04/05/2021	04/05/2021
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	04/05/2021	04/05/2021
				CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	04/05/2021	04/05/2021
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	04/05/2021	04/05/2021
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	04/05/2021	04/05/2021
				CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	04/05/2021	04/05/2021
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	04/05/2021	04/05/2021
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	04/05/2021	04/05/2021
				CT233	Surgical	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	04/05/2021	04/05/2021
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	04/05/2021	04/05/2021
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	04/05/2021	04/05/2021
				CT2406	Implantable	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	04/05/2021	04/05/2021
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	04/05/2021	04/05/2021
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	04/05/2021	04/05/2021
				CT2407	Active implantable	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	04/05/2021	04/05/2021
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	04/05/2021	04/05/2021
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	04/05/2021	04/05/2021
				CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	04/05/2021	04/05/2021
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	04/05/2021	04/05/2021
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	04/05/2021	04/05/2021
				CT2727	Active implant devices/ancillaries	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	04/05/2021	04/05/2021
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	04/05/2021	04/05/2021
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	04/05/2021	04/05/2021
				CT2750	Implantable cardiac electrical stimulators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	04/05/2021	04/05/2021
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	04/05/2021	04/05/2021

							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	04/05/2021	04/05/2021
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	04/05/2021	04/05/2021
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	04/05/2021	04/05/2021
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	04/05/2021	04/05/2021
					CT2985	Active	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	04/05/2021	04/05/2021
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	04/05/2021	04/05/2021
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	04/05/2021	04/05/2021
					CT3001	Implantable metal	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	04/05/2021	04/05/2021
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	04/05/2021	04/05/2021
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	04/05/2021	04/05/2021
					CT3002	Implantable synthetic polymer	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	04/05/2021	04/05/2021
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	04/05/2021	04/05/2021
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	04/05/2021	04/05/2021
					CT3006	Implantable inorganic material	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	04/05/2021	04/05/2021
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	04/05/2021	04/05/2021
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	04/05/2021	04/05/2021
					CT314	Electrically powered	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	04/05/2021	04/05/2021
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	04/05/2021	04/05/2021
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	04/05/2021	04/05/2021
					CT321	Long-term surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	04/05/2021	04/05/2021
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	04/05/2021	04/05/2021
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	04/05/2021	04/05/2021
					CT449	Generators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	04/05/2021	04/05/2021
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	04/05/2021	04/05/2021
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	04/05/2021	04/05/2021
					CT981	Single use	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	04/05/2021	04/05/2021
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	04/05/2021	04/05/2021
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	04/05/2021	04/05/2021
					CT983	Surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	04/05/2021	04/05/2021
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	04/05/2021	04/05/2021
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	04/05/2021	04/05/2021
10642	2021	All other reportable incidents	47265	Dual-chamber implantable pacemaker, rate-responsive	CT141	Electrodes/Leads and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A04	Material Integrity Problem	02/02/2021	02/02/2021
									A070101	Failure to Capture	02/02/2021	02/02/2021
									A23	Use of Device Problem	02/02/2021	02/02/2021
							F05	Delay to Treatment/ Therapy	A04	Material Integrity Problem	02/02/2021	02/02/2021
									A070101	Failure to Capture	02/02/2021	02/02/2021
									A23	Use of Device Problem	02/02/2021	02/02/2021
							F1905	Device Revision or Replacement	A04	Material Integrity Problem	02/02/2021	02/02/2021
									A070101	Failure to Capture	02/02/2021	02/02/2021
									A23	Use of Device Problem	02/02/2021	02/02/2021
							F1908	Prolonged Surgery	A04	Material Integrity Problem	02/02/2021	02/02/2021
									A070101	Failure to Capture	02/02/2021	02/02/2021
									A23	Use of Device Problem	02/02/2021	02/02/2021
					CT145	Cardiovascular devices	E2403	No Clinical Signs, Symptoms or Conditions	A04	Material Integrity Problem	02/02/2021	02/02/2021
									A070101	Failure to Capture	02/02/2021	02/02/2021
									A23	Use of Device Problem	02/02/2021	02/02/2021
							F05	Delay to Treatment/ Therapy	A04	Material Integrity Problem	02/02/2021	02/02/2021
									A070101	Failure to Capture	02/02/2021	02/02/2021
									A23	Use of Device Problem	02/02/2021	02/02/2021
							F1905	Device Revision or Replacement	A04	Material Integrity Problem	02/02/2021	02/02/2021
									A070101	Failure to Capture	02/02/2021	02/02/2021
									A23	Use of Device Problem	02/02/2021	02/02/2021
							F1908	Prolonged Surgery	A04	Material Integrity Problem	02/02/2021	02/02/2021
									A070101	Failure to Capture	02/02/2021	02/02/2021
									A23	Use of Device Problem	02/02/2021	02/02/2021
					CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A04	Material Integrity Problem	02/02/2021	02/02/2021
									A070101	Failure to Capture	02/02/2021	02/02/2021
									A23	Use of Device Problem	02/02/2021	02/02/2021
							F05	Delay to Treatment/ Therapy	A04	Material Integrity Problem	02/02/2021	02/02/2021
									A070101	Failure to Capture	02/02/2021	02/02/2021
									A23	Use of Device Problem	02/02/2021	02/02/2021
							F1905	Device Revision or Replacement	A04	Material Integrity Problem	02/02/2021	02/02/2021
									A070101	Failure to Capture	02/02/2021	02/02/2021
									A23	Use of Device Problem	02/02/2021	02/02/2021
							F1908	Prolonged Surgery	A04	Material Integrity Problem	02/02/2021	02/02/2021
									A070101	Failure to Capture	02/02/2021	02/02/2021
									A23	Use of Device Problem	02/02/2021	02/02/2021
					CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A04	Material Integrity Problem	02/02/2021	02/02/2021
									A070101	Failure to Capture	02/02/2021	02/02/2021
									A23	Use of Device Problem	02/02/2021	02/02/2021
							F05	Delay to Treatment/ Therapy	A04	Material Integrity Problem	02/02/2021	02/02/2021
									A070101	Failure to Capture	02/02/2021	02/02/2021
									A23	Use of Device Problem	02/02/2021	02/02/2021
							F1905	Device Revision or Replacement	A04	Material Integrity Problem	02/02/2021	02/02/2021
									A070101	Failure to Capture	02/02/2021	02/02/2021
									A23	Use of Device Problem	02/02/2021	02/02/2021
							F1908	Prolonged Surgery	A04	Material Integrity Problem	02/02/2021	02/02/2021
									A070101	Failure to Capture	02/02/2021	02/02/2021
									A23	Use of Device Problem	02/02/2021	02/02/2021
					CT233	Surgical	E2403	No Clinical Signs, Symptoms or Conditions	A04	Material Integrity Problem	02/02/2021	02/02/2021
									A070101	Failure to Capture	02/02/2021	02/02/2021
									A23	Use of Device Problem	02/02/2021	02/02/2021
							F05	Delay to Treatment/ Therapy	A04	Material Integrity Problem	02/02/2021	02/02/2021
									A070101	Failure to Capture	02/02/2021	02/02/2021
									A23	Use of Device Problem	02/02/2021	02/02/2021
							F1905	Device Revision or Replacement	A04	Material Integrity Problem	02/02/2021	02/02/2021
									A070101	Failure to Capture	02/02/2021	02/02/2021
									A23	Use of Device Problem	02/02/2021	02/02/2021

		F1908	Prolonged Surgery	A04	Material Integrity Problem	02/02/2021	02/02/2021
				A070101	Failure to Capture	02/02/2021	02/02/2021
				A23	Use of Device Problem	02/02/2021	02/02/2021
CT2406	Implantable	E2403	No Clinical Signs, Symptoms or Conditions	A04	Material Integrity Problem	02/02/2021	02/02/2021
				A070101	Failure to Capture	02/02/2021	02/02/2021
				A23	Use of Device Problem	02/02/2021	02/02/2021
		F05	Delay to Treatment/ Therapy	A04	Material Integrity Problem	02/02/2021	02/02/2021
				A070101	Failure to Capture	02/02/2021	02/02/2021
				A23	Use of Device Problem	02/02/2021	02/02/2021
		F1905	Device Revision or Replacement	A04	Material Integrity Problem	02/02/2021	02/02/2021
				A070101	Failure to Capture	02/02/2021	02/02/2021
				A23	Use of Device Problem	02/02/2021	02/02/2021
		F1908	Prolonged Surgery	A04	Material Integrity Problem	02/02/2021	02/02/2021
				A070101	Failure to Capture	02/02/2021	02/02/2021
				A23	Use of Device Problem	02/02/2021	02/02/2021
CT2407	Active implantable	E2403	No Clinical Signs, Symptoms or Conditions	A04	Material Integrity Problem	02/02/2021	02/02/2021
				A070101	Failure to Capture	02/02/2021	02/02/2021
				A23	Use of Device Problem	02/02/2021	02/02/2021
		F05	Delay to Treatment/ Therapy	A04	Material Integrity Problem	02/02/2021	02/02/2021
				A070101	Failure to Capture	02/02/2021	02/02/2021
				A23	Use of Device Problem	02/02/2021	02/02/2021
		F1905	Device Revision or Replacement	A04	Material Integrity Problem	02/02/2021	02/02/2021
				A070101	Failure to Capture	02/02/2021	02/02/2021
				A23	Use of Device Problem	02/02/2021	02/02/2021
		F1908	Prolonged Surgery	A04	Material Integrity Problem	02/02/2021	02/02/2021
				A070101	Failure to Capture	02/02/2021	02/02/2021
				A23	Use of Device Problem	02/02/2021	02/02/2021
CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A04	Material Integrity Problem	02/02/2021	02/02/2021
				A070101	Failure to Capture	02/02/2021	02/02/2021
				A23	Use of Device Problem	02/02/2021	02/02/2021
		F05	Delay to Treatment/ Therapy	A04	Material Integrity Problem	02/02/2021	02/02/2021
				A070101	Failure to Capture	02/02/2021	02/02/2021
				A23	Use of Device Problem	02/02/2021	02/02/2021
		F1905	Device Revision or Replacement	A04	Material Integrity Problem	02/02/2021	02/02/2021
				A070101	Failure to Capture	02/02/2021	02/02/2021
				A23	Use of Device Problem	02/02/2021	02/02/2021
		F1908	Prolonged Surgery	A04	Material Integrity Problem	02/02/2021	02/02/2021
				A070101	Failure to Capture	02/02/2021	02/02/2021
				A23	Use of Device Problem	02/02/2021	02/02/2021
CT2727	Active implant devices/ancillaries	E2403	No Clinical Signs, Symptoms or Conditions	A04	Material Integrity Problem	02/02/2021	02/02/2021
				A070101	Failure to Capture	02/02/2021	02/02/2021
				A23	Use of Device Problem	02/02/2021	02/02/2021
		F05	Delay to Treatment/ Therapy	A04	Material Integrity Problem	02/02/2021	02/02/2021
				A070101	Failure to Capture	02/02/2021	02/02/2021
				A23	Use of Device Problem	02/02/2021	02/02/2021
		F1905	Device Revision or Replacement	A04	Material Integrity Problem	02/02/2021	02/02/2021
				A070101	Failure to Capture	02/02/2021	02/02/2021
				A23	Use of Device Problem	02/02/2021	02/02/2021
		F1908	Prolonged Surgery	A04	Material Integrity Problem	02/02/2021	02/02/2021
				A070101	Failure to Capture	02/02/2021	02/02/2021
				A23	Use of Device Problem	02/02/2021	02/02/2021
CT2750	Implantable cardiac electrical stimulators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A04	Material Integrity Problem	02/02/2021	02/02/2021
				A070101	Failure to Capture	02/02/2021	02/02/2021
				A23	Use of Device Problem	02/02/2021	02/02/2021
		F05	Delay to Treatment/ Therapy	A04	Material Integrity Problem	02/02/2021	02/02/2021
				A070101	Failure to Capture	02/02/2021	02/02/2021
				A23	Use of Device Problem	02/02/2021	02/02/2021
		F1905	Device Revision or Replacement	A04	Material Integrity Problem	02/02/2021	02/02/2021
				A070101	Failure to Capture	02/02/2021	02/02/2021
				A23	Use of Device Problem	02/02/2021	02/02/2021
		F1908	Prolonged Surgery	A04	Material Integrity Problem	02/02/2021	02/02/2021
				A070101	Failure to Capture	02/02/2021	02/02/2021
				A23	Use of Device Problem	02/02/2021	02/02/2021
CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A04	Material Integrity Problem	02/02/2021	02/02/2021
				A070101	Failure to Capture	02/02/2021	02/02/2021
				A23	Use of Device Problem	02/02/2021	02/02/2021
		F05	Delay to Treatment/ Therapy	A04	Material Integrity Problem	02/02/2021	02/02/2021
				A070101	Failure to Capture	02/02/2021	02/02/2021
				A23	Use of Device Problem	02/02/2021	02/02/2021
		F1905	Device Revision or Replacement	A04	Material Integrity Problem	02/02/2021	02/02/2021
				A070101	Failure to Capture	02/02/2021	02/02/2021
				A23	Use of Device Problem	02/02/2021	02/02/2021
		F1908	Prolonged Surgery	A04	Material Integrity Problem	02/02/2021	02/02/2021
				A070101	Failure to Capture	02/02/2021	02/02/2021
				A23	Use of Device Problem	02/02/2021	02/02/2021
CT2985	Active	E2403	No Clinical Signs, Symptoms or Conditions	A04	Material Integrity Problem	02/02/2021	02/02/2021
				A070101	Failure to Capture	02/02/2021	02/02/2021
				A23	Use of Device Problem	02/02/2021	02/02/2021
		F05	Delay to Treatment/ Therapy	A04	Material Integrity Problem	02/02/2021	02/02/2021
				A070101	Failure to Capture	02/02/2021	02/02/2021
				A23	Use of Device Problem	02/02/2021	02/02/2021
		F1905	Device Revision or Replacement	A04	Material Integrity Problem	02/02/2021	02/02/2021
				A070101	Failure to Capture	02/02/2021	02/02/2021
				A23	Use of Device Problem	02/02/2021	02/02/2021
		F1908	Prolonged Surgery	A04	Material Integrity Problem	02/02/2021	02/02/2021
				A070101	Failure to Capture	02/02/2021	02/02/2021
				A23	Use of Device Problem	02/02/2021	02/02/2021
CT3001	Implantable metal	E2403	No Clinical Signs, Symptoms or Conditions	A04	Material Integrity Problem	02/02/2021	02/02/2021

								A070101	Failure to Capture	02/02/2021	02/02/2021
								A23	Use of Device Problem	02/02/2021	02/02/2021
						F05	Delay to Treatment/ Therapy	A04	Material Integrity Problem	02/02/2021	02/02/2021
								A070101	Failure to Capture	02/02/2021	02/02/2021
								A23	Use of Device Problem	02/02/2021	02/02/2021
						F1905	Device Revision or Replacement	A04	Material Integrity Problem	02/02/2021	02/02/2021
								A070101	Failure to Capture	02/02/2021	02/02/2021
								A23	Use of Device Problem	02/02/2021	02/02/2021
						F1908	Prolonged Surgery	A04	Material Integrity Problem	02/02/2021	02/02/2021
								A070101	Failure to Capture	02/02/2021	02/02/2021
								A23	Use of Device Problem	02/02/2021	02/02/2021
					CT3002	Implantable synthetic polymer	E2403	No Clinical Signs, Symptoms or Conditions	A04	Material Integrity Problem	02/02/2021	02/02/2021
									A070101	Failure to Capture	02/02/2021	02/02/2021
									A23	Use of Device Problem	02/02/2021	02/02/2021
							F05	Delay to Treatment/ Therapy	A04	Material Integrity Problem	02/02/2021	02/02/2021
									A070101	Failure to Capture	02/02/2021	02/02/2021
									A23	Use of Device Problem	02/02/2021	02/02/2021
							F1905	Device Revision or Replacement	A04	Material Integrity Problem	02/02/2021	02/02/2021
									A070101	Failure to Capture	02/02/2021	02/02/2021
									A23	Use of Device Problem	02/02/2021	02/02/2021
							F1908	Prolonged Surgery	A04	Material Integrity Problem	02/02/2021	02/02/2021
									A070101	Failure to Capture	02/02/2021	02/02/2021
									A23	Use of Device Problem	02/02/2021	02/02/2021
					CT3006	Implantable inorganic material	E2403	No Clinical Signs, Symptoms or Conditions	A04	Material Integrity Problem	02/02/2021	02/02/2021
									A070101	Failure to Capture	02/02/2021	02/02/2021
									A23	Use of Device Problem	02/02/2021	02/02/2021
							F05	Delay to Treatment/ Therapy	A04	Material Integrity Problem	02/02/2021	02/02/2021
									A070101	Failure to Capture	02/02/2021	02/02/2021
									A23	Use of Device Problem	02/02/2021	02/02/2021
							F1905	Device Revision or Replacement	A04	Material Integrity Problem	02/02/2021	02/02/2021
									A070101	Failure to Capture	02/02/2021	02/02/2021
									A23	Use of Device Problem	02/02/2021	02/02/2021
							F1908	Prolonged Surgery	A04	Material Integrity Problem	02/02/2021	02/02/2021
									A070101	Failure to Capture	02/02/2021	02/02/2021
									A23	Use of Device Problem	02/02/2021	02/02/2021
					CT314	Electrically powered	E2403	No Clinical Signs, Symptoms or Conditions	A04	Material Integrity Problem	02/02/2021	02/02/2021
									A070101	Failure to Capture	02/02/2021	02/02/2021
									A23	Use of Device Problem	02/02/2021	02/02/2021
							F05	Delay to Treatment/ Therapy	A04	Material Integrity Problem	02/02/2021	02/02/2021
									A070101	Failure to Capture	02/02/2021	02/02/2021
									A23	Use of Device Problem	02/02/2021	02/02/2021
							F1905	Device Revision or Replacement	A04	Material Integrity Problem	02/02/2021	02/02/2021
									A070101	Failure to Capture	02/02/2021	02/02/2021
									A23	Use of Device Problem	02/02/2021	02/02/2021
							F1908	Prolonged Surgery	A04	Material Integrity Problem	02/02/2021	02/02/2021
									A070101	Failure to Capture	02/02/2021	02/02/2021
									A23	Use of Device Problem	02/02/2021	02/02/2021
					CT321	Long-term surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A04	Material Integrity Problem	02/02/2021	02/02/2021
									A070101	Failure to Capture	02/02/2021	02/02/2021
									A23	Use of Device Problem	02/02/2021	02/02/2021
							F05	Delay to Treatment/ Therapy	A04	Material Integrity Problem	02/02/2021	02/02/2021
									A070101	Failure to Capture	02/02/2021	02/02/2021
									A23	Use of Device Problem	02/02/2021	02/02/2021
							F1905	Device Revision or Replacement	A04	Material Integrity Problem	02/02/2021	02/02/2021
									A070101	Failure to Capture	02/02/2021	02/02/2021
									A23	Use of Device Problem	02/02/2021	02/02/2021
							F1908	Prolonged Surgery	A04	Material Integrity Problem	02/02/2021	02/02/2021
									A070101	Failure to Capture	02/02/2021	02/02/2021
									A23	Use of Device Problem	02/02/2021	02/02/2021
					CT449	Generators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A04	Material Integrity Problem	02/02/2021	02/02/2021
									A070101	Failure to Capture	02/02/2021	02/02/2021
									A23	Use of Device Problem	02/02/2021	02/02/2021
							F05	Delay to Treatment/ Therapy	A04	Material Integrity Problem	02/02/2021	02/02/2021
									A070101	Failure to Capture	02/02/2021	02/02/2021
									A23	Use of Device Problem	02/02/2021	02/02/2021
							F1905	Device Revision or Replacement	A04	Material Integrity Problem	02/02/2021	02/02/2021
									A070101	Failure to Capture	02/02/2021	02/02/2021
									A23	Use of Device Problem	02/02/2021	02/02/2021
							F1908	Prolonged Surgery	A04	Material Integrity Problem	02/02/2021	02/02/2021
									A070101	Failure to Capture	02/02/2021	02/02/2021
									A23	Use of Device Problem	02/02/2021	02/02/2021
					CT981	Single use	E2403	No Clinical Signs, Symptoms or Conditions	A04	Material Integrity Problem	02/02/2021	02/02/2021
									A070101	Failure to Capture	02/02/2021	02/02/2021
									A23	Use of Device Problem	02/02/2021	02/02/2021
							F05	Delay to Treatment/ Therapy	A04	Material Integrity Problem	02/02/2021	02/02/2021
									A070101	Failure to Capture	02/02/2021	02/02/2021
									A23	Use of Device Problem	02/02/2021	02/02/2021
							F1905	Device Revision or Replacement	A04	Material Integrity Problem	02/02/2021	02/02/2021
									A070101	Failure to Capture	02/02/2021	02/02/2021
									A23	Use of Device Problem	02/02/2021	02/02/2021
							F1908	Prolonged Surgery	A04	Material Integrity Problem	02/02/2021	02/02/2021
									A070101	Failure to Capture	02/02/2021	02/02/2021
									A23	Use of Device Problem	02/02/2021	02/02/2021
					CT983	Surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A04	Material Integrity Problem	02/02/2021	02/02/2021
									A070101	Failure to Capture	02/02/2021	02/02/2021
									A23	Use of Device Problem	02/02/2021	02/02/2021
							F05	Delay to Treatment/ Therapy	A04	Material Integrity Problem	02/02/2021	02/02/2021
									A070101	Failure to Capture	02/02/2021	02/02/2021

									A23	Use of Device Problem	02/02/2021	02/02/2021
							F1905	Device Revision or Replacement	A04	Material Integrity Problem	02/02/2021	02/02/2021
									A070101	Failure to Capture	02/02/2021	02/02/2021
									A23	Use of Device Problem	02/02/2021	02/02/2021
							F1908	Prolonged Surgery	A04	Material Integrity Problem	02/02/2021	02/02/2021
									A070101	Failure to Capture	02/02/2021	02/02/2021
									A23	Use of Device Problem	02/02/2021	02/02/2021
10643	2021	All other reportable incidents	35223	Endocardial/interventricular septal pacing lead	CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A07	Electrical /Electronic Property Problem	26/04/2021	26/04/2021
									A070102	High Capture Threshold	26/04/2021	26/04/2021
									A070901	Decreased Sensitivity	26/04/2021	26/04/2021
									A072202	Low Impedance	26/04/2021	26/04/2021
									A1205	Loose or Intermittent Connection	26/04/2021	26/04/2021
									A180103	Device Contamination with Body Fluid	26/04/2021	26/04/2021
							F26	No Health Consequences or Impact	A07	Electrical /Electronic Property Problem	26/04/2021	26/04/2021
									A070102	High Capture Threshold	26/04/2021	26/04/2021
									A070901	Decreased Sensitivity	26/04/2021	26/04/2021
									A072202	Low Impedance	26/04/2021	26/04/2021
									A1205	Loose or Intermittent Connection	26/04/2021	26/04/2021
									A180103	Device Contamination with Body Fluid	26/04/2021	26/04/2021
10644	2021	All other reportable incidents	35852	Single-chamber implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A070909	Oversensing		
									A071302	Inappropriate/Inadequate Shock/Stimulation		
									A090801	Signal Artifact/Noise		
									A1301	Failure to Read Input Signal		
							F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A070909	Oversensing		
									A071302	Inappropriate/Inadequate Shock/Stimulation		
									A090801	Signal Artifact/Noise		
									A1301	Failure to Read Input Signal		
							F26	No Health Consequences or Impact	A070909	Oversensing		
									A071302	Inappropriate/Inadequate Shock/Stimulation		
									A090801	Signal Artifact/Noise		
									A1301	Failure to Read Input Signal		
10645	2021	All other reportable incidents	47263	Cardiac resynchronization therapy implantable pacemaker	CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A090804	False Positive Result		
							F26	No Health Consequences or Impact	A090804	False Positive Result		
					CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A090804	False Positive Result		
							F26	No Health Consequences or Impact	A090804	False Positive Result		
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A090804	False Positive Result		
							F26	No Health Consequences or Impact	A090804	False Positive Result		
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A090804	False Positive Result		
							F26	No Health Consequences or Impact	A090804	False Positive Result		
10646	2021	All other reportable incidents	47270	Cardiac resynchronization therapy implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A070910	Undersensing		
									A0712	Pacing Problem		
							F08	Hospitalization or Prolonged Hospitalization	A070910	Undersensing		
									A0712	Pacing Problem		
10647	2021	All other reportable incidents	35852	Single-chamber implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem	22/02/2021	22/02/2021
									A070504	Premature Discharge of Battery	22/02/2021	22/02/2021
									A070602	Delayed Charge Time	22/02/2021	22/02/2021
							F08	Hospitalization or Prolonged Hospitalization	A0705	Battery Problem	22/02/2021	22/02/2021
									A070504	Premature Discharge of Battery	22/02/2021	22/02/2021
									A070602	Delayed Charge Time	22/02/2021	22/02/2021
							F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A0705	Battery Problem	22/02/2021	22/02/2021
									A070504	Premature Discharge of Battery	22/02/2021	22/02/2021
									A070602	Delayed Charge Time	22/02/2021	22/02/2021
							F1905	Device Revision or Replacement	A0705	Battery Problem	22/02/2021	22/02/2021
									A070504	Premature Discharge of Battery	22/02/2021	22/02/2021
									A070602	Delayed Charge Time	22/02/2021	22/02/2021
10648	2021	All other reportable incidents	47265	Dual-chamber implantable pacemaker, rate-responsive	CT141	Electrodes/Leads and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	01/02/2021	01/02/2021
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	01/02/2021	01/02/2021
							F12	Serious Injury/ Illness/ Impairment	A070504	Premature Discharge of Battery	01/02/2021	01/02/2021
							F19	Surgical Intervention	A070504	Premature Discharge of Battery	01/02/2021	01/02/2021
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	01/02/2021	01/02/2021
					CT145	Cardiovascular devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	01/02/2021	01/02/2021
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	01/02/2021	01/02/2021
							F12	Serious Injury/ Illness/ Impairment	A070504	Premature Discharge of Battery	01/02/2021	01/02/2021
							F19	Surgical Intervention	A070504	Premature Discharge of Battery	01/02/2021	01/02/2021
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	01/02/2021	01/02/2021
					CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	01/02/2021	01/02/2021
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	01/02/2021	01/02/2021
							F12	Serious Injury/ Illness/ Impairment	A070504	Premature Discharge of Battery	01/02/2021	01/02/2021
							F19	Surgical Intervention	A070504	Premature Discharge of Battery	01/02/2021	01/02/2021
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	01/02/2021	01/02/2021
					CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	01/02/2021	01/02/2021
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	01/02/2021	01/02/2021
							F12	Serious Injury/ Illness/ Impairment	A070504	Premature Discharge of Battery	01/02/2021	01/02/2021
							F19	Surgical Intervention	A070504	Premature Discharge of Battery	01/02/2021	01/02/2021
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	01/02/2021	01/02/2021
					CT233	Surgical	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	01/02/2021	01/02/2021
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	01/02/2021	01/02/2021
							F12	Serious Injury/ Illness/ Impairment	A070504	Premature Discharge of Battery	01/02/2021	01/02/2021
							F19	Surgical Intervention	A070504	Premature Discharge of Battery	01/02/2021	01/02/2021
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	01/02/2021	01/02/2021
					CT2406	Implantable	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	01/02/2021	01/02/2021
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	01/02/2021	01/02/2021
							F12	Serious Injury/ Illness/ Impairment	A070504	Premature Discharge of Battery	01/02/2021	01/02/2021
							F19	Surgical Intervention	A070504	Premature Discharge of Battery	01/02/2021	01/02/2021
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	01/02/2021	01/02/2021
					CT2407	Active implantable	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	01/02/2021	01/02/2021
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	01/02/2021	01/02/2021
							F12	Serious Injury/ Illness/ Impairment	A070504	Premature Discharge of Battery	01/02/2021	01/02/2021

							F19	Surgical Intervention	A070504	Premature Discharge of Battery	01/02/2021	01/02/2021
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	01/02/2021	01/02/2021
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	01/02/2021	01/02/2021
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	01/02/2021	01/02/2021
							F12	Serious Injury/ Illness/ Impairment	A070504	Premature Discharge of Battery	01/02/2021	01/02/2021
							F19	Surgical Intervention	A070504	Premature Discharge of Battery	01/02/2021	01/02/2021
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	01/02/2021	01/02/2021
					CT2727	Active implant devices/ancillaries	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	01/02/2021	01/02/2021
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	01/02/2021	01/02/2021
							F12	Serious Injury/ Illness/ Impairment	A070504	Premature Discharge of Battery	01/02/2021	01/02/2021
							F19	Surgical Intervention	A070504	Premature Discharge of Battery	01/02/2021	01/02/2021
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	01/02/2021	01/02/2021
					CT2750	Implantable cardiac electrical stimulators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	01/02/2021	01/02/2021
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	01/02/2021	01/02/2021
							F12	Serious Injury/ Illness/ Impairment	A070504	Premature Discharge of Battery	01/02/2021	01/02/2021
							F19	Surgical Intervention	A070504	Premature Discharge of Battery	01/02/2021	01/02/2021
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	01/02/2021	01/02/2021
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	01/02/2021	01/02/2021
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	01/02/2021	01/02/2021
							F12	Serious Injury/ Illness/ Impairment	A070504	Premature Discharge of Battery	01/02/2021	01/02/2021
							F19	Surgical Intervention	A070504	Premature Discharge of Battery	01/02/2021	01/02/2021
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	01/02/2021	01/02/2021
					CT2985	Active	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	01/02/2021	01/02/2021
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	01/02/2021	01/02/2021
							F12	Serious Injury/ Illness/ Impairment	A070504	Premature Discharge of Battery	01/02/2021	01/02/2021
							F19	Surgical Intervention	A070504	Premature Discharge of Battery	01/02/2021	01/02/2021
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	01/02/2021	01/02/2021
					CT3001	Implantable metal	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	01/02/2021	01/02/2021
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	01/02/2021	01/02/2021
							F12	Serious Injury/ Illness/ Impairment	A070504	Premature Discharge of Battery	01/02/2021	01/02/2021
							F19	Surgical Intervention	A070504	Premature Discharge of Battery	01/02/2021	01/02/2021
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	01/02/2021	01/02/2021
					CT3002	Implantable synthetic polymer	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	01/02/2021	01/02/2021
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	01/02/2021	01/02/2021
							F12	Serious Injury/ Illness/ Impairment	A070504	Premature Discharge of Battery	01/02/2021	01/02/2021
							F19	Surgical Intervention	A070504	Premature Discharge of Battery	01/02/2021	01/02/2021
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	01/02/2021	01/02/2021
					CT3006	Implantable inorganic material	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	01/02/2021	01/02/2021
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	01/02/2021	01/02/2021
							F12	Serious Injury/ Illness/ Impairment	A070504	Premature Discharge of Battery	01/02/2021	01/02/2021
							F19	Surgical Intervention	A070504	Premature Discharge of Battery	01/02/2021	01/02/2021
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	01/02/2021	01/02/2021
					CT314	Electrically powered	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	01/02/2021	01/02/2021
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	01/02/2021	01/02/2021
							F12	Serious Injury/ Illness/ Impairment	A070504	Premature Discharge of Battery	01/02/2021	01/02/2021
							F19	Surgical Intervention	A070504	Premature Discharge of Battery	01/02/2021	01/02/2021
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	01/02/2021	01/02/2021
					CT321	Long-term surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	01/02/2021	01/02/2021
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	01/02/2021	01/02/2021
							F12	Serious Injury/ Illness/ Impairment	A070504	Premature Discharge of Battery	01/02/2021	01/02/2021
							F19	Surgical Intervention	A070504	Premature Discharge of Battery	01/02/2021	01/02/2021
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	01/02/2021	01/02/2021
					CT449	Generators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	01/02/2021	01/02/2021
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	01/02/2021	01/02/2021
							F12	Serious Injury/ Illness/ Impairment	A070504	Premature Discharge of Battery	01/02/2021	01/02/2021
							F19	Surgical Intervention	A070504	Premature Discharge of Battery	01/02/2021	01/02/2021
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	01/02/2021	01/02/2021
					CT981	Single use	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	01/02/2021	01/02/2021
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	01/02/2021	01/02/2021
							F12	Serious Injury/ Illness/ Impairment	A070504	Premature Discharge of Battery	01/02/2021	01/02/2021
							F19	Surgical Intervention	A070504	Premature Discharge of Battery	01/02/2021	01/02/2021
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	01/02/2021	01/02/2021
					CT983	Surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	01/02/2021	01/02/2021
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	01/02/2021	01/02/2021
							F12	Serious Injury/ Illness/ Impairment	A070504	Premature Discharge of Battery	01/02/2021	01/02/2021
							F19	Surgical Intervention	A070504	Premature Discharge of Battery	01/02/2021	01/02/2021
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	01/02/2021	01/02/2021
10649	2021	All other reportable incidents	47263	Cardiac resynchronization therapy implantable pacemaker	CT1684	Cardiac pacemakers and associated devices	E2401	Insufficient Information	A090804	False Positive Result		
							E2403	No Clinical Signs, Symptoms or Conditions	A090804	False Positive Result		
							F24	Insufficient Health Impact Information	A090804	False Positive Result		
					CT1685	Cardiac pacemakers	E2401	Insufficient Information	A090804	False Positive Result		
							E2403	No Clinical Signs, Symptoms or Conditions	A090804	False Positive Result		
							F24	Insufficient Health Impact Information	A090804	False Positive Result		
					CT257	Implantable cardiac pacemakers	E2401	Insufficient Information	A090804	False Positive Result		
							E2403	No Clinical Signs, Symptoms or Conditions	A090804	False Positive Result		
							F24	Insufficient Health Impact Information	A090804	False Positive Result		
					CT2866	Implantable cardiac electrical stimulators	E2401	Insufficient Information	A090804	False Positive Result		
							E2403	No Clinical Signs, Symptoms or Conditions	A090804	False Positive Result		
							F24	Insufficient Health Impact Information	A090804	False Positive Result		
10650	2021	All other reportable incidents	47265	Dual-chamber implantable pacemaker, rate-responsive	CT141	Electrodes/Leads and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	16/03/2021	16/03/2021
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	16/03/2021	16/03/2021
							F12	Serious Injury/ Illness/ Impairment	A070504	Premature Discharge of Battery	16/03/2021	16/03/2021
							F1903	Device Explantation	A070504	Premature Discharge of Battery	16/03/2021	16/03/2021
					CT145	Cardiovascular devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	16/03/2021	16/03/2021
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	16/03/2021	16/03/2021
							F12	Serious Injury/ Illness/ Impairment	A070504	Premature Discharge of Battery	16/03/2021	16/03/2021
							F1903	Device Explantation	A070504	Premature Discharge of Battery	16/03/2021	16/03/2021
					CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	16/03/2021	16/03/2021

							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	16/03/2021	16/03/2021
							F12	Serious Injury/ Illness/ Impairment	A070504	Premature Discharge of Battery	16/03/2021	16/03/2021
							F1903	Device Explantation	A070504	Premature Discharge of Battery	16/03/2021	16/03/2021
					CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	16/03/2021	16/03/2021
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	16/03/2021	16/03/2021
							F12	Serious Injury/ Illness/ Impairment	A070504	Premature Discharge of Battery	16/03/2021	16/03/2021
							F1903	Device Explantation	A070504	Premature Discharge of Battery	16/03/2021	16/03/2021
					CT233	Surgical	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	16/03/2021	16/03/2021
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	16/03/2021	16/03/2021
							F12	Serious Injury/ Illness/ Impairment	A070504	Premature Discharge of Battery	16/03/2021	16/03/2021
							F1903	Device Explantation	A070504	Premature Discharge of Battery	16/03/2021	16/03/2021
					CT2406	Implantable	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	16/03/2021	16/03/2021
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	16/03/2021	16/03/2021
							F12	Serious Injury/ Illness/ Impairment	A070504	Premature Discharge of Battery	16/03/2021	16/03/2021
							F1903	Device Explantation	A070504	Premature Discharge of Battery	16/03/2021	16/03/2021
					CT2407	Active implantable	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	16/03/2021	16/03/2021
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	16/03/2021	16/03/2021
							F12	Serious Injury/ Illness/ Impairment	A070504	Premature Discharge of Battery	16/03/2021	16/03/2021
							F1903	Device Explantation	A070504	Premature Discharge of Battery	16/03/2021	16/03/2021
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	16/03/2021	16/03/2021
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	16/03/2021	16/03/2021
							F12	Serious Injury/ Illness/ Impairment	A070504	Premature Discharge of Battery	16/03/2021	16/03/2021
							F1903	Device Explantation	A070504	Premature Discharge of Battery	16/03/2021	16/03/2021
					CT2727	Active implant devices/ancillaries	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	16/03/2021	16/03/2021
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	16/03/2021	16/03/2021
							F12	Serious Injury/ Illness/ Impairment	A070504	Premature Discharge of Battery	16/03/2021	16/03/2021
							F1903	Device Explantation	A070504	Premature Discharge of Battery	16/03/2021	16/03/2021
					CT2750	Implantable cardiac electrical stimulators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	16/03/2021	16/03/2021
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	16/03/2021	16/03/2021
							F12	Serious Injury/ Illness/ Impairment	A070504	Premature Discharge of Battery	16/03/2021	16/03/2021
							F1903	Device Explantation	A070504	Premature Discharge of Battery	16/03/2021	16/03/2021
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	16/03/2021	16/03/2021
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	16/03/2021	16/03/2021
							F12	Serious Injury/ Illness/ Impairment	A070504	Premature Discharge of Battery	16/03/2021	16/03/2021
							F1903	Device Explantation	A070504	Premature Discharge of Battery	16/03/2021	16/03/2021
					CT2985	Active	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	16/03/2021	16/03/2021
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	16/03/2021	16/03/2021
							F12	Serious Injury/ Illness/ Impairment	A070504	Premature Discharge of Battery	16/03/2021	16/03/2021
							F1903	Device Explantation	A070504	Premature Discharge of Battery	16/03/2021	16/03/2021
					CT3001	Implantable metal	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	16/03/2021	16/03/2021
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	16/03/2021	16/03/2021
							F12	Serious Injury/ Illness/ Impairment	A070504	Premature Discharge of Battery	16/03/2021	16/03/2021
							F1903	Device Explantation	A070504	Premature Discharge of Battery	16/03/2021	16/03/2021
					CT3002	Implantable synthetic polymer	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	16/03/2021	16/03/2021
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	16/03/2021	16/03/2021
							F12	Serious Injury/ Illness/ Impairment	A070504	Premature Discharge of Battery	16/03/2021	16/03/2021
							F1903	Device Explantation	A070504	Premature Discharge of Battery	16/03/2021	16/03/2021
					CT3006	Implantable inorganic material	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	16/03/2021	16/03/2021
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	16/03/2021	16/03/2021
							F12	Serious Injury/ Illness/ Impairment	A070504	Premature Discharge of Battery	16/03/2021	16/03/2021
							F1903	Device Explantation	A070504	Premature Discharge of Battery	16/03/2021	16/03/2021
					CT314	Electrically powered	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	16/03/2021	16/03/2021
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	16/03/2021	16/03/2021
							F12	Serious Injury/ Illness/ Impairment	A070504	Premature Discharge of Battery	16/03/2021	16/03/2021
							F1903	Device Explantation	A070504	Premature Discharge of Battery	16/03/2021	16/03/2021
					CT321	Long-term surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	16/03/2021	16/03/2021
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	16/03/2021	16/03/2021
							F12	Serious Injury/ Illness/ Impairment	A070504	Premature Discharge of Battery	16/03/2021	16/03/2021
							F1903	Device Explantation	A070504	Premature Discharge of Battery	16/03/2021	16/03/2021
					CT449	Generators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	16/03/2021	16/03/2021
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	16/03/2021	16/03/2021
							F12	Serious Injury/ Illness/ Impairment	A070504	Premature Discharge of Battery	16/03/2021	16/03/2021
							F1903	Device Explantation	A070504	Premature Discharge of Battery	16/03/2021	16/03/2021
					CT981	Single use	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	16/03/2021	16/03/2021
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	16/03/2021	16/03/2021
							F12	Serious Injury/ Illness/ Impairment	A070504	Premature Discharge of Battery	16/03/2021	16/03/2021
							F1903	Device Explantation	A070504	Premature Discharge of Battery	16/03/2021	16/03/2021
					CT983	Surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	16/03/2021	16/03/2021
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	16/03/2021	16/03/2021
							F12	Serious Injury/ Illness/ Impairment	A070504	Premature Discharge of Battery	16/03/2021	16/03/2021
							F1903	Device Explantation	A070504	Premature Discharge of Battery	16/03/2021	16/03/2021
10651	2021	All other reportable incidents	35852	Single-chamber implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	09/02/2021	09/02/2021
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	09/02/2021	09/02/2021
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	09/02/2021	09/02/2021
10652	2021	All other reportable incidents	17882	Physiologic-monitoring defibrillation system	CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A0719	Unexpected Shutdown		
							F26	No Health Consequences or Impact	A0719	Unexpected Shutdown		
10653	2021	All other reportable incidents	35852	Single-chamber implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A072201	High Impedance		
							F26	No Health Consequences or Impact	A072201	High Impedance		
10654	2021	All other reportable incidents	47267	Single-chamber implantable pacemaker, rate-responsive	CT141	Electrodes/Leads and associated devices	E2401	Insufficient Information	A070504	Premature Discharge of Battery	04/03/2021	04/03/2021
							E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	04/03/2021	04/03/2021
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	04/03/2021	04/03/2021
							F12	Serious Injury/ Illness/ Impairment	A070504	Premature Discharge of Battery	04/03/2021	04/03/2021
							F19	Surgical Intervention	A070504	Premature Discharge of Battery	04/03/2021	04/03/2021
					CT145	Cardiovascular devices	E2401	Insufficient Information	A070504	Premature Discharge of Battery	04/03/2021	04/03/2021
							E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	04/03/2021	04/03/2021
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	04/03/2021	04/03/2021
							F12	Serious Injury/ Illness/ Impairment	A070504	Premature Discharge of Battery	04/03/2021	04/03/2021
							F19	Surgical Intervention	A070504	Premature Discharge of Battery	04/03/2021	04/03/2021

					CT1684	Cardiac pacemakers and associated devices	E2401	Insufficient Information	A070504	Premature Discharge of Battery	04/03/2021	04/03/2021
							E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	04/03/2021	04/03/2021
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	04/03/2021	04/03/2021
							F12	Serious Injury/ Illness/ Impairment	A070504	Premature Discharge of Battery	04/03/2021	04/03/2021
							F19	Surgical Intervention	A070504	Premature Discharge of Battery	04/03/2021	04/03/2021
					CT1685	Cardiac pacemakers	E2401	Insufficient Information	A070504	Premature Discharge of Battery	04/03/2021	04/03/2021
							E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	04/03/2021	04/03/2021
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	04/03/2021	04/03/2021
							F12	Serious Injury/ Illness/ Impairment	A070504	Premature Discharge of Battery	04/03/2021	04/03/2021
							F19	Surgical Intervention	A070504	Premature Discharge of Battery	04/03/2021	04/03/2021
					CT233	Surgical	E2401	Insufficient Information	A070504	Premature Discharge of Battery	04/03/2021	04/03/2021
							E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	04/03/2021	04/03/2021
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	04/03/2021	04/03/2021
							F12	Serious Injury/ Illness/ Impairment	A070504	Premature Discharge of Battery	04/03/2021	04/03/2021
							F19	Surgical Intervention	A070504	Premature Discharge of Battery	04/03/2021	04/03/2021
					CT2406	Implantable	E2401	Insufficient Information	A070504	Premature Discharge of Battery	04/03/2021	04/03/2021
							E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	04/03/2021	04/03/2021
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	04/03/2021	04/03/2021
							F12	Serious Injury/ Illness/ Impairment	A070504	Premature Discharge of Battery	04/03/2021	04/03/2021
							F19	Surgical Intervention	A070504	Premature Discharge of Battery	04/03/2021	04/03/2021
					CT2407	Active implantable	E2401	Insufficient Information	A070504	Premature Discharge of Battery	04/03/2021	04/03/2021
							E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	04/03/2021	04/03/2021
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	04/03/2021	04/03/2021
							F12	Serious Injury/ Illness/ Impairment	A070504	Premature Discharge of Battery	04/03/2021	04/03/2021
							F19	Surgical Intervention	A070504	Premature Discharge of Battery	04/03/2021	04/03/2021
					CT257	Implantable cardiac pacemakers	E2401	Insufficient Information	A070504	Premature Discharge of Battery	04/03/2021	04/03/2021
							E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	04/03/2021	04/03/2021
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	04/03/2021	04/03/2021
							F12	Serious Injury/ Illness/ Impairment	A070504	Premature Discharge of Battery	04/03/2021	04/03/2021
							F19	Surgical Intervention	A070504	Premature Discharge of Battery	04/03/2021	04/03/2021
					CT2727	Active implant devices/ancillaries	E2401	Insufficient Information	A070504	Premature Discharge of Battery	04/03/2021	04/03/2021
							E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	04/03/2021	04/03/2021
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	04/03/2021	04/03/2021
							F12	Serious Injury/ Illness/ Impairment	A070504	Premature Discharge of Battery	04/03/2021	04/03/2021
							F19	Surgical Intervention	A070504	Premature Discharge of Battery	04/03/2021	04/03/2021
					CT2750	Implantable cardiac electrical stimulators and associated devices	E2401	Insufficient Information	A070504	Premature Discharge of Battery	04/03/2021	04/03/2021
							E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	04/03/2021	04/03/2021
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	04/03/2021	04/03/2021
							F12	Serious Injury/ Illness/ Impairment	A070504	Premature Discharge of Battery	04/03/2021	04/03/2021
							F19	Surgical Intervention	A070504	Premature Discharge of Battery	04/03/2021	04/03/2021
					CT2866	Implantable cardiac electrical stimulators	E2401	Insufficient Information	A070504	Premature Discharge of Battery	04/03/2021	04/03/2021
							E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	04/03/2021	04/03/2021
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	04/03/2021	04/03/2021
							F12	Serious Injury/ Illness/ Impairment	A070504	Premature Discharge of Battery	04/03/2021	04/03/2021
							F19	Surgical Intervention	A070504	Premature Discharge of Battery	04/03/2021	04/03/2021
					CT2985	Active	E2401	Insufficient Information	A070504	Premature Discharge of Battery	04/03/2021	04/03/2021
							E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	04/03/2021	04/03/2021
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	04/03/2021	04/03/2021
							F12	Serious Injury/ Illness/ Impairment	A070504	Premature Discharge of Battery	04/03/2021	04/03/2021
							F19	Surgical Intervention	A070504	Premature Discharge of Battery	04/03/2021	04/03/2021
					CT3001	Implantable metal	E2401	Insufficient Information	A070504	Premature Discharge of Battery	04/03/2021	04/03/2021
							E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	04/03/2021	04/03/2021
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	04/03/2021	04/03/2021
							F12	Serious Injury/ Illness/ Impairment	A070504	Premature Discharge of Battery	04/03/2021	04/03/2021
							F19	Surgical Intervention	A070504	Premature Discharge of Battery	04/03/2021	04/03/2021
					CT3002	Implantable synthetic polymer	E2401	Insufficient Information	A070504	Premature Discharge of Battery	04/03/2021	04/03/2021
							E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	04/03/2021	04/03/2021
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	04/03/2021	04/03/2021
							F12	Serious Injury/ Illness/ Impairment	A070504	Premature Discharge of Battery	04/03/2021	04/03/2021
							F19	Surgical Intervention	A070504	Premature Discharge of Battery	04/03/2021	04/03/2021
					CT3006	Implantable inorganic material	E2401	Insufficient Information	A070504	Premature Discharge of Battery	04/03/2021	04/03/2021
							E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	04/03/2021	04/03/2021
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	04/03/2021	04/03/2021
							F12	Serious Injury/ Illness/ Impairment	A070504	Premature Discharge of Battery	04/03/2021	04/03/2021
							F19	Surgical Intervention	A070504	Premature Discharge of Battery	04/03/2021	04/03/2021
					CT314	Electrically powered	E2401	Insufficient Information	A070504	Premature Discharge of Battery	04/03/2021	04/03/2021
							E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	04/03/2021	04/03/2021
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	04/03/2021	04/03/2021
							F12	Serious Injury/ Illness/ Impairment	A070504	Premature Discharge of Battery	04/03/2021	04/03/2021
							F19	Surgical Intervention	A070504	Premature Discharge of Battery	04/03/2021	04/03/2021
					CT321	Long-term surgical invasive	E2401	Insufficient Information	A070504	Premature Discharge of Battery	04/03/2021	04/03/2021
							E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	04/03/2021	04/03/2021
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	04/03/2021	04/03/2021
							F12	Serious Injury/ Illness/ Impairment	A070504	Premature Discharge of Battery	04/03/2021	04/03/2021
							F19	Surgical Intervention	A070504	Premature Discharge of Battery	04/03/2021	04/03/2021
					CT449	Generators and associated devices	E2401	Insufficient Information	A070504	Premature Discharge of Battery	04/03/2021	04/03/2021
							E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	04/03/2021	04/03/2021
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	04/03/2021	04/03/2021
							F12	Serious Injury/ Illness/ Impairment	A070504	Premature Discharge of Battery	04/03/2021	04/03/2021
							F19	Surgical Intervention	A070504	Premature Discharge of Battery	04/03/2021	04/03/2021
					CT981	Single use	E2401	Insufficient Information	A070504	Premature Discharge of Battery	04/03/2021	04/03/2021
							E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	04/03/2021	04/03/2021
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	04/03/2021	04/03/2021
							F12	Serious Injury/ Illness/ Impairment	A070504	Premature Discharge of Battery	04/03/2021	04/03/2021
							F19	Surgical Intervention	A070504	Premature Discharge of Battery	04/03/2021	04/03/2021
					CT983	Surgical invasive	E2401	Insufficient Information	A070504	Premature Discharge of Battery	04/03/2021	04/03/2021
							E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	04/03/2021	04/03/2021
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	04/03/2021	04/03/2021

							F12	Serious Injury/ Illness/ Impairment	A070504	Premature Discharge of Battery	04/03/2021	04/03/2021
							F19	Surgical Intervention	A070504	Premature Discharge of Battery	04/03/2021	04/03/2021
10655	2021	All other reportable incidents	37265	Dual-chamber implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	29/04/2021	29/04/2021
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	29/04/2021	29/04/2021
							F1903	Device Explantation	A070504	Premature Discharge of Battery	29/04/2021	29/04/2021
10656	2021	All other reportable incidents	36241	Pacemaker/defibrillator lead	CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A040101	Fracture		
									A072201	High Impedance		
									A090801	Signal Artifact/Noise		
							F19	Surgical Intervention	A040101	Fracture		
									A072201	High Impedance		
									A090801	Signal Artifact/Noise		
							F1905	Device Revision or Replacement	A040101	Fracture		
									A072201	High Impedance		
									A090801	Signal Artifact/Noise		
10657	2021	All other reportable incidents	47270	Cardiac resynchronization therapy implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A070909	Oversensing		
									A072202	Low Impedance		
							F26	No Health Consequences or Impact	A070909	Oversensing		
									A072202	Low Impedance		
10658	2021	All other reportable incidents	35223	Endocardial/interventricular septal pacing lead	CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070909	Oversensing		
									A090801	Signal Artifact/Noise		
							F26	No Health Consequences or Impact	A070909	Oversensing		
									A090801	Signal Artifact/Noise		
10659	2021	All other reportable incidents	35223	Endocardial/interventricular septal pacing lead	CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A0401	Break		
									A070909	Oversensing		
									A090801	Signal Artifact/Noise		
							F26	No Health Consequences or Impact	A0401	Break		
									A070909	Oversensing		
									A090801	Signal Artifact/Noise		
10660	2021	All other reportable incidents	35852	Single-chamber implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E2104	Electric Shock	A070504	Premature Discharge of Battery	14/04/2021	14/04/2021
									A070909	Oversensing	14/04/2021	14/04/2021
									A071302	Inappropriate/Inadequate Shock/Stimulation	14/04/2021	14/04/2021
							F1903	Device Explantation	A070504	Premature Discharge of Battery	14/04/2021	14/04/2021
									A070909	Oversensing	14/04/2021	14/04/2021
									A071302	Inappropriate/Inadequate Shock/Stimulation	14/04/2021	14/04/2021
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	14/04/2021	14/04/2021
									A070909	Oversensing	14/04/2021	14/04/2021
									A071302	Inappropriate/Inadequate Shock/Stimulation	14/04/2021	14/04/2021
10661	2021	All other reportable incidents	37265	Dual-chamber implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E2401	Insufficient Information	A160602	Premature End-of-Life Indicator		
							F14	Prolonged Episode of Care	A160602	Premature End-of-Life Indicator		
							F24	Insufficient Health Impact Information	A160602	Premature End-of-Life Indicator		
10662	2021	All other reportable incidents	37265	Dual-chamber implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	12/03/2021	12/03/2021
							F19	Surgical Intervention	A070504	Premature Discharge of Battery	12/03/2021	12/03/2021
10663	2021	All other reportable incidents	36241	Pacemaker/defibrillator lead	CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070101	Failure to Capture		
							F26	No Health Consequences or Impact	A070101	Failure to Capture		
10664	2021	All other reportable incidents	35852	Single-chamber implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E2401	Insufficient Information	A070504	Premature Discharge of Battery		
									A0908	Incorrect, Inadequate or Imprecise Result or Readings		
							E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
									A0908	Incorrect, Inadequate or Imprecise Result or Readings		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
									A0908	Incorrect, Inadequate or Imprecise Result or Readings		
10665	2021	All other reportable incidents	37265	Dual-chamber implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	05/03/2021	05/03/2021
									A0908	Incorrect, Inadequate or Imprecise Result or Readings	05/03/2021	05/03/2021
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	05/03/2021	05/03/2021
									A0908	Incorrect, Inadequate or Imprecise Result or Readings	05/03/2021	05/03/2021
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	05/03/2021	05/03/2021
									A0908	Incorrect, Inadequate or Imprecise Result or Readings	05/03/2021	05/03/2021
10666	2021	All other reportable incidents	47267	Single-chamber implantable pacemaker, rate-responsive	CT141	Electrodes/Leads and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem	26/03/2021	26/03/2021
									A070504	Premature Discharge of Battery	26/03/2021	26/03/2021
							F08	Hospitalization or Prolonged Hospitalization	A0705	Battery Problem	26/03/2021	26/03/2021
									A070504	Premature Discharge of Battery	26/03/2021	26/03/2021
							F12	Serious Injury/ Illness/ Impairment	A0705	Battery Problem	26/03/2021	26/03/2021
									A070504	Premature Discharge of Battery	26/03/2021	26/03/2021
							F1903	Device Explantation	A0705	Battery Problem	26/03/2021	26/03/2021
									A070504	Premature Discharge of Battery	26/03/2021	26/03/2021
					CT145	Cardiovascular devices	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem	26/03/2021	26/03/2021
									A070504	Premature Discharge of Battery	26/03/2021	26/03/2021
							F08	Hospitalization or Prolonged Hospitalization	A0705	Battery Problem	26/03/2021	26/03/2021
									A070504	Premature Discharge of Battery	26/03/2021	26/03/2021
							F12	Serious Injury/ Illness/ Impairment	A0705	Battery Problem	26/03/2021	26/03/2021
									A070504	Premature Discharge of Battery	26/03/2021	26/03/2021
							F1903	Device Explantation	A0705	Battery Problem	26/03/2021	26/03/2021
									A070504	Premature Discharge of Battery	26/03/2021	26/03/2021
					CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem	26/03/2021	26/03/2021
									A070504	Premature Discharge of Battery	26/03/2021	26/03/2021
							F08	Hospitalization or Prolonged Hospitalization	A0705	Battery Problem	26/03/2021	26/03/2021
									A070504	Premature Discharge of Battery	26/03/2021	26/03/2021
							F12	Serious Injury/ Illness/ Impairment	A0705	Battery Problem	26/03/2021	26/03/2021
									A070504	Premature Discharge of Battery	26/03/2021	26/03/2021
							F1903	Device Explantation	A0705	Battery Problem	26/03/2021	26/03/2021
									A070504	Premature Discharge of Battery	26/03/2021	26/03/2021
					CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem	26/03/2021	26/03/2021
									A070504	Premature Discharge of Battery	26/03/2021	26/03/2021
							F08	Hospitalization or Prolonged Hospitalization	A0705	Battery Problem	26/03/2021	26/03/2021

								A070504	Premature Discharge of Battery	26/03/2021	26/03/2021
						F12	Serious Injury/ Illness/ Impairment	A0705	Battery Problem	26/03/2021	26/03/2021
								A070504	Premature Discharge of Battery	26/03/2021	26/03/2021
						F1903	Device Explantation	A0705	Battery Problem	26/03/2021	26/03/2021
								A070504	Premature Discharge of Battery	26/03/2021	26/03/2021
					CT233	Surgical	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem	26/03/2021	26/03/2021
									A070504	Premature Discharge of Battery	26/03/2021	26/03/2021
							F08	Hospitalization or Prolonged Hospitalization	A0705	Battery Problem	26/03/2021	26/03/2021
									A070504	Premature Discharge of Battery	26/03/2021	26/03/2021
							F12	Serious Injury/ Illness/ Impairment	A0705	Battery Problem	26/03/2021	26/03/2021
									A070504	Premature Discharge of Battery	26/03/2021	26/03/2021
							F1903	Device Explantation	A0705	Battery Problem	26/03/2021	26/03/2021
									A070504	Premature Discharge of Battery	26/03/2021	26/03/2021
					CT2406	Implantable	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem	26/03/2021	26/03/2021
									A070504	Premature Discharge of Battery	26/03/2021	26/03/2021
							F08	Hospitalization or Prolonged Hospitalization	A0705	Battery Problem	26/03/2021	26/03/2021
									A070504	Premature Discharge of Battery	26/03/2021	26/03/2021
							F12	Serious Injury/ Illness/ Impairment	A0705	Battery Problem	26/03/2021	26/03/2021
									A070504	Premature Discharge of Battery	26/03/2021	26/03/2021
							F1903	Device Explantation	A0705	Battery Problem	26/03/2021	26/03/2021
									A070504	Premature Discharge of Battery	26/03/2021	26/03/2021
					CT2407	Active implantable	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem	26/03/2021	26/03/2021
									A070504	Premature Discharge of Battery	26/03/2021	26/03/2021
							F08	Hospitalization or Prolonged Hospitalization	A0705	Battery Problem	26/03/2021	26/03/2021
									A070504	Premature Discharge of Battery	26/03/2021	26/03/2021
							F12	Serious Injury/ Illness/ Impairment	A0705	Battery Problem	26/03/2021	26/03/2021
									A070504	Premature Discharge of Battery	26/03/2021	26/03/2021
							F1903	Device Explantation	A0705	Battery Problem	26/03/2021	26/03/2021
									A070504	Premature Discharge of Battery	26/03/2021	26/03/2021
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem	26/03/2021	26/03/2021
									A070504	Premature Discharge of Battery	26/03/2021	26/03/2021
							F08	Hospitalization or Prolonged Hospitalization	A0705	Battery Problem	26/03/2021	26/03/2021
									A070504	Premature Discharge of Battery	26/03/2021	26/03/2021
							F12	Serious Injury/ Illness/ Impairment	A0705	Battery Problem	26/03/2021	26/03/2021
									A070504	Premature Discharge of Battery	26/03/2021	26/03/2021
							F1903	Device Explantation	A0705	Battery Problem	26/03/2021	26/03/2021
									A070504	Premature Discharge of Battery	26/03/2021	26/03/2021
					CT2727	Active implant devices/ancillaries	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem	26/03/2021	26/03/2021
									A070504	Premature Discharge of Battery	26/03/2021	26/03/2021
							F08	Hospitalization or Prolonged Hospitalization	A0705	Battery Problem	26/03/2021	26/03/2021
									A070504	Premature Discharge of Battery	26/03/2021	26/03/2021
							F12	Serious Injury/ Illness/ Impairment	A0705	Battery Problem	26/03/2021	26/03/2021
									A070504	Premature Discharge of Battery	26/03/2021	26/03/2021
							F1903	Device Explantation	A0705	Battery Problem	26/03/2021	26/03/2021
									A070504	Premature Discharge of Battery	26/03/2021	26/03/2021
					CT2750	Implantable cardiac electrical stimulators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem	26/03/2021	26/03/2021
									A070504	Premature Discharge of Battery	26/03/2021	26/03/2021
							F08	Hospitalization or Prolonged Hospitalization	A0705	Battery Problem	26/03/2021	26/03/2021
									A070504	Premature Discharge of Battery	26/03/2021	26/03/2021
							F12	Serious Injury/ Illness/ Impairment	A0705	Battery Problem	26/03/2021	26/03/2021
									A070504	Premature Discharge of Battery	26/03/2021	26/03/2021
							F1903	Device Explantation	A0705	Battery Problem	26/03/2021	26/03/2021
									A070504	Premature Discharge of Battery	26/03/2021	26/03/2021
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem	26/03/2021	26/03/2021
									A070504	Premature Discharge of Battery	26/03/2021	26/03/2021
							F08	Hospitalization or Prolonged Hospitalization	A0705	Battery Problem	26/03/2021	26/03/2021
									A070504	Premature Discharge of Battery	26/03/2021	26/03/2021
							F12	Serious Injury/ Illness/ Impairment	A0705	Battery Problem	26/03/2021	26/03/2021
									A070504	Premature Discharge of Battery	26/03/2021	26/03/2021
							F1903	Device Explantation	A0705	Battery Problem	26/03/2021	26/03/2021
									A070504	Premature Discharge of Battery	26/03/2021	26/03/2021
					CT2985	Active	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem	26/03/2021	26/03/2021
									A070504	Premature Discharge of Battery	26/03/2021	26/03/2021
							F08	Hospitalization or Prolonged Hospitalization	A0705	Battery Problem	26/03/2021	26/03/2021
									A070504	Premature Discharge of Battery	26/03/2021	26/03/2021
							F12	Serious Injury/ Illness/ Impairment	A0705	Battery Problem	26/03/2021	26/03/2021
									A070504	Premature Discharge of Battery	26/03/2021	26/03/2021
							F1903	Device Explantation	A0705	Battery Problem	26/03/2021	26/03/2021
									A070504	Premature Discharge of Battery	26/03/2021	26/03/2021
					CT3001	Implantable metal	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem	26/03/2021	26/03/2021
									A070504	Premature Discharge of Battery	26/03/2021	26/03/2021
							F08	Hospitalization or Prolonged Hospitalization	A0705	Battery Problem	26/03/2021	26/03/2021
									A070504	Premature Discharge of Battery	26/03/2021	26/03/2021
							F12	Serious Injury/ Illness/ Impairment	A0705	Battery Problem	26/03/2021	26/03/2021
									A070504	Premature Discharge of Battery	26/03/2021	26/03/2021
							F1903	Device Explantation	A0705	Battery Problem	26/03/2021	26/03/2021
									A070504	Premature Discharge of Battery	26/03/2021	26/03/2021
					CT3002	Implantable synthetic polymer	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem	26/03/2021	26/03/2021
									A070504	Premature Discharge of Battery	26/03/2021	26/03/2021
							F08	Hospitalization or Prolonged Hospitalization	A0705	Battery Problem	26/03/2021	26/03/2021
									A070504	Premature Discharge of Battery	26/03/2021	26/03/2021
							F12	Serious Injury/ Illness/ Impairment	A0705	Battery Problem	26/03/2021	26/03/2021
									A070504	Premature Discharge of Battery	26/03/2021	26/03/2021
							F1903	Device Explantation	A0705	Battery Problem	26/03/2021	26/03/2021
									A070504	Premature Discharge of Battery	26/03/2021	26/03/2021
					CT3006	Implantable inorganic material	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem	26/03/2021	26/03/2021
									A070504	Premature Discharge of Battery	26/03/2021	26/03/2021
							F08	Hospitalization or Prolonged Hospitalization	A0705	Battery Problem	26/03/2021	26/03/2021

									A070504	Premature Discharge of Battery	26/03/2021	26/03/2021
							F12	Serious Injury/ Illness/ Impairment	A0705	Battery Problem	26/03/2021	26/03/2021
									A070504	Premature Discharge of Battery	26/03/2021	26/03/2021
							F1903	Device Explantation	A0705	Battery Problem	26/03/2021	26/03/2021
									A070504	Premature Discharge of Battery	26/03/2021	26/03/2021
					CT314	Electrically powered	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem	26/03/2021	26/03/2021
									A070504	Premature Discharge of Battery	26/03/2021	26/03/2021
							F08	Hospitalization or Prolonged Hospitalization	A0705	Battery Problem	26/03/2021	26/03/2021
									A070504	Premature Discharge of Battery	26/03/2021	26/03/2021
							F12	Serious Injury/ Illness/ Impairment	A0705	Battery Problem	26/03/2021	26/03/2021
									A070504	Premature Discharge of Battery	26/03/2021	26/03/2021
							F1903	Device Explantation	A0705	Battery Problem	26/03/2021	26/03/2021
									A070504	Premature Discharge of Battery	26/03/2021	26/03/2021
					CT321	Long-term surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem	26/03/2021	26/03/2021
									A070504	Premature Discharge of Battery	26/03/2021	26/03/2021
							F08	Hospitalization or Prolonged Hospitalization	A0705	Battery Problem	26/03/2021	26/03/2021
									A070504	Premature Discharge of Battery	26/03/2021	26/03/2021
							F12	Serious Injury/ Illness/ Impairment	A0705	Battery Problem	26/03/2021	26/03/2021
									A070504	Premature Discharge of Battery	26/03/2021	26/03/2021
							F1903	Device Explantation	A0705	Battery Problem	26/03/2021	26/03/2021
									A070504	Premature Discharge of Battery	26/03/2021	26/03/2021
					CT449	Generators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem	26/03/2021	26/03/2021
									A070504	Premature Discharge of Battery	26/03/2021	26/03/2021
							F08	Hospitalization or Prolonged Hospitalization	A0705	Battery Problem	26/03/2021	26/03/2021
									A070504	Premature Discharge of Battery	26/03/2021	26/03/2021
							F12	Serious Injury/ Illness/ Impairment	A0705	Battery Problem	26/03/2021	26/03/2021
									A070504	Premature Discharge of Battery	26/03/2021	26/03/2021
							F1903	Device Explantation	A0705	Battery Problem	26/03/2021	26/03/2021
									A070504	Premature Discharge of Battery	26/03/2021	26/03/2021
					CT981	Single use	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem	26/03/2021	26/03/2021
									A070504	Premature Discharge of Battery	26/03/2021	26/03/2021
							F08	Hospitalization or Prolonged Hospitalization	A0705	Battery Problem	26/03/2021	26/03/2021
									A070504	Premature Discharge of Battery	26/03/2021	26/03/2021
							F12	Serious Injury/ Illness/ Impairment	A0705	Battery Problem	26/03/2021	26/03/2021
									A070504	Premature Discharge of Battery	26/03/2021	26/03/2021
							F1903	Device Explantation	A0705	Battery Problem	26/03/2021	26/03/2021
									A070504	Premature Discharge of Battery	26/03/2021	26/03/2021
					CT983	Surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem	26/03/2021	26/03/2021
									A070504	Premature Discharge of Battery	26/03/2021	26/03/2021
							F08	Hospitalization or Prolonged Hospitalization	A0705	Battery Problem	26/03/2021	26/03/2021
									A070504	Premature Discharge of Battery	26/03/2021	26/03/2021
							F12	Serious Injury/ Illness/ Impairment	A0705	Battery Problem	26/03/2021	26/03/2021
									A070504	Premature Discharge of Battery	26/03/2021	26/03/2021
							F1903	Device Explantation	A0705	Battery Problem	26/03/2021	26/03/2021
									A070504	Premature Discharge of Battery	26/03/2021	26/03/2021
10667	2021	All other reportable incidents	37265	Dual-chamber implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E060109	Tachycardia	A071201	Failure to Convert Rhythm		
							E2312	Fatigue	A071201	Failure to Convert Rhythm		
							E2330	Pain	A071201	Failure to Convert Rhythm		
							E233001	Chest Pain	A071201	Failure to Convert Rhythm		
							F08	Hospitalization or Prolonged Hospitalization	A071201	Failure to Convert Rhythm		
							F14	Prolonged Episode of Care	A071201	Failure to Convert Rhythm		
							F19	Surgical Intervention	A071201	Failure to Convert Rhythm		
10668	2021	All other reportable incidents	47265	Dual-chamber implantable pacemaker, rate-responsive	CT141	Electrodes/Leads and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT145	Cardiovascular devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT233	Surgical	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT2406	Implantable	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT2407	Active implantable	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT2727	Active implant devices/ancillaries	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT2750	Implantable cardiac electrical stimulators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT2985	Active	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		

					CT3001	Implantable metal	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT3002	Implantable synthetic polymer	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT3006	Implantable inorganic material	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT314	Electrically powered	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT321	Long-term surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT449	Generators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT981	Single use	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT983	Surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
10669	2021	All other reportable incidents	36241	Pacemaker/defibrillator lead	CT1684	Cardiac pacemakers and associated devices	E2401	Insufficient Information	A072201	High Impedance		
							E2403	No Clinical Signs, Symptoms or Conditions	A072201	High Impedance		
							F24	Insufficient Health Impact Information	A072201	High Impedance		
10670	2021	All other reportable incidents	35852	Single-chamber implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A072201	High Impedance		
							F26	No Health Consequences or Impact	A072201	High Impedance		
10671	2021	All other reportable incidents	47263	Cardiac resynchronization therapy implantable pacemaker	CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
10672	2021	All other reportable incidents	47265	Dual-chamber implantable pacemaker, rate-responsive	CT141	Electrodes/Leads and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	02/03/2021	02/03/2021
							F19	Surgical Intervention	A070504	Premature Discharge of Battery	02/03/2021	02/03/2021
					CT145	Cardiovascular devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	02/03/2021	02/03/2021
							F19	Surgical Intervention	A070504	Premature Discharge of Battery	02/03/2021	02/03/2021
					CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	02/03/2021	02/03/2021
							F19	Surgical Intervention	A070504	Premature Discharge of Battery	02/03/2021	02/03/2021
					CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	02/03/2021	02/03/2021
							F19	Surgical Intervention	A070504	Premature Discharge of Battery	02/03/2021	02/03/2021
					CT233	Surgical	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	02/03/2021	02/03/2021
							F19	Surgical Intervention	A070504	Premature Discharge of Battery	02/03/2021	02/03/2021
					CT2406	Implantable	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	02/03/2021	02/03/2021
							F19	Surgical Intervention	A070504	Premature Discharge of Battery	02/03/2021	02/03/2021
					CT2407	Active implantable	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	02/03/2021	02/03/2021
							F19	Surgical Intervention	A070504	Premature Discharge of Battery	02/03/2021	02/03/2021
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	02/03/2021	02/03/2021
							F19	Surgical Intervention	A070504	Premature Discharge of Battery	02/03/2021	02/03/2021
					CT2727	Active implant devices/ancillaries	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	02/03/2021	02/03/2021
							F19	Surgical Intervention	A070504	Premature Discharge of Battery	02/03/2021	02/03/2021
					CT2750	Implantable cardiac electrical stimulators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	02/03/2021	02/03/2021
							F19	Surgical Intervention	A070504	Premature Discharge of Battery	02/03/2021	02/03/2021
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	02/03/2021	02/03/2021
							F19	Surgical Intervention	A070504	Premature Discharge of Battery	02/03/2021	02/03/2021
					CT2985	Active	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	02/03/2021	02/03/2021
							F19	Surgical Intervention	A070504	Premature Discharge of Battery	02/03/2021	02/03/2021
					CT3001	Implantable metal	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	02/03/2021	02/03/2021
							F19	Surgical Intervention	A070504	Premature Discharge of Battery	02/03/2021	02/03/2021
					CT3002	Implantable synthetic polymer	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	02/03/2021	02/03/2021
							F19	Surgical Intervention	A070504	Premature Discharge of Battery	02/03/2021	02/03/2021
					CT3006	Implantable inorganic material	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	02/03/2021	02/03/2021
							F19	Surgical Intervention	A070504	Premature Discharge of Battery	02/03/2021	02/03/2021
					CT314	Electrically powered	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	02/03/2021	02/03/2021
							F19	Surgical Intervention	A070504	Premature Discharge of Battery	02/03/2021	02/03/2021
					CT321	Long-term surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	02/03/2021	02/03/2021
							F19	Surgical Intervention	A070504	Premature Discharge of Battery	02/03/2021	02/03/2021
					CT449	Generators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	02/03/2021	02/03/2021
							F19	Surgical Intervention	A070504	Premature Discharge of Battery	02/03/2021	02/03/2021
					CT981	Single use	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	02/03/2021	02/03/2021
							F19	Surgical Intervention	A070504	Premature Discharge of Battery	02/03/2021	02/03/2021
					CT983	Surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	02/03/2021	02/03/2021
							F19	Surgical Intervention	A070504	Premature Discharge of Battery	02/03/2021	02/03/2021
10673	2021	All other reportable incidents	17882	Physiologic-monitoring defibrillation system	CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070803	Failure to Power Up		
							F26	No Health Consequences or Impact	A070803	Failure to Power Up		
10674	2021	All other reportable incidents	17882	Physiologic-monitoring defibrillation system	CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070803	Failure to Power Up		
							F27	Problem Identified During Non-Clinical Procedure	A070803	Failure to Power Up		
10675	2021	All other reportable incidents	17882	Physiologic-monitoring defibrillation system	CT1684	Cardiac pacemakers and associated devices	E2401	Insufficient Information	A070803	Failure to Power Up		
							E2403	No Clinical Signs, Symptoms or Conditions	A070803	Failure to Power Up		
							F1001	Absence of Treatment	A070803	Failure to Power Up		
							F26	No Health Consequences or Impact	A070803	Failure to Power Up		
10676	2021	All other reportable incidents	17882	Physiologic-monitoring defibrillation system	CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A0719	Unexpected Shutdown		
							F27	Problem Identified During Non-Clinical Procedure	A0719	Unexpected Shutdown		
10677	2021	All other reportable incidents	47270	Cardiac resynchronization therapy implantable	CT2866	Implantable cardiac electrical stimulators	E2401	Insufficient Information	A070909	Oversensing		

				defibrillator					A0712	Pacing Problem		
							E2403	No Clinical Signs, Symptoms or Conditions	A070909	Oversensing		
									A0712	Pacing Problem		
							F26	No Health Consequences or Impact	A070909	Oversensing		
									A0712	Pacing Problem		
10678	2021	All other reportable incidents	37265	Dual-chamber implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A070603	Failure to Charge	09/03/2021	09/03/2021
							F1901	Additional Surgery	A070603	Failure to Charge	09/03/2021	09/03/2021
							F1905	Device Revision or Replacement	A070603	Failure to Charge	09/03/2021	09/03/2021
10679	2021	All other reportable incidents	36241	Pacemaker/defibrillator lead	CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A0203	Defective Device		
									A070102	High Capture Threshold		
									A070909	Oversensing		
									A072202	Low Impedance		
									A090801	Signal Artifact/Noise		
							F24	Insufficient Health Impact Information	A0203	Defective Device		
									A070102	High Capture Threshold		
									A070909	Oversensing		
									A072202	Low Impedance		
									A090801	Signal Artifact/Noise		
10680	2021	All other reportable incidents	47265	Dual-chamber implantable pacemaker, rate-responsive	CT141	Electrodes/Leads and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	29/07/2021	29/07/2021
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	29/07/2021	29/07/2021
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	29/07/2021	29/07/2021
					CT145	Cardiovascular devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	29/07/2021	29/07/2021
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	29/07/2021	29/07/2021
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	29/07/2021	29/07/2021
					CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	29/07/2021	29/07/2021
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	29/07/2021	29/07/2021
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	29/07/2021	29/07/2021
					CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	29/07/2021	29/07/2021
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	29/07/2021	29/07/2021
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	29/07/2021	29/07/2021
					CT233	Surgical	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	29/07/2021	29/07/2021
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	29/07/2021	29/07/2021
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	29/07/2021	29/07/2021
					CT2406	Implantable	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	29/07/2021	29/07/2021
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	29/07/2021	29/07/2021
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	29/07/2021	29/07/2021
					CT2407	Active implantable	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	29/07/2021	29/07/2021
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	29/07/2021	29/07/2021
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	29/07/2021	29/07/2021
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	29/07/2021	29/07/2021
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	29/07/2021	29/07/2021
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	29/07/2021	29/07/2021
					CT2727	Active implant devices/ancillaries	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	29/07/2021	29/07/2021
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	29/07/2021	29/07/2021
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	29/07/2021	29/07/2021
					CT2750	Implantable cardiac electrical stimulators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	29/07/2021	29/07/2021
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	29/07/2021	29/07/2021
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	29/07/2021	29/07/2021
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	29/07/2021	29/07/2021
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	29/07/2021	29/07/2021
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	29/07/2021	29/07/2021
					CT2985	Active	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	29/07/2021	29/07/2021
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	29/07/2021	29/07/2021
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	29/07/2021	29/07/2021
					CT3001	Implantable metal	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	29/07/2021	29/07/2021
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	29/07/2021	29/07/2021
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	29/07/2021	29/07/2021
					CT3002	Implantable synthetic polymer	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	29/07/2021	29/07/2021
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	29/07/2021	29/07/2021
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	29/07/2021	29/07/2021
					CT3006	Implantable inorganic material	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	29/07/2021	29/07/2021
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	29/07/2021	29/07/2021
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	29/07/2021	29/07/2021
					CT314	Electrically powered	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	29/07/2021	29/07/2021
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	29/07/2021	29/07/2021
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	29/07/2021	29/07/2021
					CT321	Long-term surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	29/07/2021	29/07/2021
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	29/07/2021	29/07/2021
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	29/07/2021	29/07/2021
					CT449	Generators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	29/07/2021	29/07/2021
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	29/07/2021	29/07/2021
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	29/07/2021	29/07/2021
					CT981	Single use	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	29/07/2021	29/07/2021
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	29/07/2021	29/07/2021
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	29/07/2021	29/07/2021
					CT983	Surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	29/07/2021	29/07/2021
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	29/07/2021	29/07/2021
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	29/07/2021	29/07/2021
10681	2021	All other reportable incidents	47265	Dual-chamber implantable pacemaker, rate-responsive	CT141	Electrodes/Leads and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT145	Cardiovascular devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		

							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT233	Surgical	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT2406	Implantable	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT2407	Active implantable	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT2727	Active implant devices/ancillaries	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT2750	Implantable cardiac electrical stimulators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT2985	Active	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT3001	Implantable metal	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT3002	Implantable synthetic polymer	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT3006	Implantable inorganic material	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT314	Electrically powered	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT321	Long-term surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT449	Generators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT981	Single use	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT983	Surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
10682	2021	All other reportable incidents	36241	Pacemaker/defibrillator lead	CT1684	Cardiac pacemakers and associated devices	E0116	Headache	A070101	Failure to Capture		
							E060104	Bradycardia	A070101	Failure to Capture		
							F1905	Device Revision or Replacement	A070101	Failure to Capture		
							F26	No Health Consequences or Impact	A070101	Failure to Capture		
10683	2021	All other reportable incidents	37265	Dual-chamber implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E060101	Asystole	A0712	Pacing Problem	05/02/2021	05/02/2021
							F19	Surgical Intervention	A0712	Pacing Problem	05/02/2021	05/02/2021
							F1905	Device Revision or Replacement	A0712	Pacing Problem	05/02/2021	05/02/2021
10684	2021	All other reportable incidents	35852	Single-chamber implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E2104	Electric Shock	A071302	Inappropriate/Inadequate Shock/Stimulation		
							E2403	No Clinical Signs, Symptoms or Conditions	A071302	Inappropriate/Inadequate Shock/Stimulation		
							F26	No Health Consequences or Impact	A071302	Inappropriate/Inadequate Shock/Stimulation		
10685	2021	All other reportable incidents	35852	Single-chamber implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem	23/03/2021	23/03/2021
									A070504	Premature Discharge of Battery	23/03/2021	23/03/2021
							F19	Surgical Intervention	A0705	Battery Problem	23/03/2021	23/03/2021
									A070504	Premature Discharge of Battery	23/03/2021	23/03/2021
10686	2021	All other reportable incidents	47205	Cardiac pulse generator programmer	CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A11	Computer Software Problem		
							F26	No Health Consequences or Impact	A11	Computer Software Problem		
10687	2021	All other reportable incidents	35223	Endocardial/interventricular septal pacing lead	CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070101	Failure to Capture		
									A090801	Signal Artifact/Noise		
							F26	No Health Consequences or Impact	A070101	Failure to Capture		
									A090801	Signal Artifact/Noise		
10688	2021	All other reportable incidents	47265	Dual-chamber implantable pacemaker, rate-responsive	CT141	Electrodes/Leads and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F27	Problem Identified During Non-Clinical Procedure	A070504	Premature Discharge of Battery		
					CT145	Cardiovascular devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F27	Problem Identified During Non-Clinical Procedure	A070504	Premature Discharge of Battery		
					CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F27	Problem Identified During Non-Clinical Procedure	A070504	Premature Discharge of Battery		
					CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F27	Problem Identified During Non-Clinical Procedure	A070504	Premature Discharge of Battery		
					CT233	Surgical	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F27	Problem Identified During Non-Clinical Procedure	A070504	Premature Discharge of Battery		
					CT2406	Implantable	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F27	Problem Identified During Non-Clinical Procedure	A070504	Premature Discharge of Battery		
					CT2407	Active implantable	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F27	Problem Identified During Non-Clinical Procedure	A070504	Premature Discharge of Battery		
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F27	Problem Identified During Non-Clinical Procedure	A070504	Premature Discharge of Battery		
					CT2727	Active implant devices/ancillaries	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F27	Problem Identified During Non-Clinical Procedure	A070504	Premature Discharge of Battery		

					CT2750	Implantable cardiac electrical stimulators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F27	Problem Identified During Non-Clinical Procedure	A070504	Premature Discharge of Battery		
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F27	Problem Identified During Non-Clinical Procedure	A070504	Premature Discharge of Battery		
					CT2985	Active	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F27	Problem Identified During Non-Clinical Procedure	A070504	Premature Discharge of Battery		
					CT3001	Implantable metal	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F27	Problem Identified During Non-Clinical Procedure	A070504	Premature Discharge of Battery		
					CT3002	Implantable synthetic polymer	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F27	Problem Identified During Non-Clinical Procedure	A070504	Premature Discharge of Battery		
					CT3006	Implantable inorganic material	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F27	Problem Identified During Non-Clinical Procedure	A070504	Premature Discharge of Battery		
					CT314	Electrically powered	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F27	Problem Identified During Non-Clinical Procedure	A070504	Premature Discharge of Battery		
					CT321	Long-term surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F27	Problem Identified During Non-Clinical Procedure	A070504	Premature Discharge of Battery		
					CT449	Generators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F27	Problem Identified During Non-Clinical Procedure	A070504	Premature Discharge of Battery		
					CT981	Single use	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F27	Problem Identified During Non-Clinical Procedure	A070504	Premature Discharge of Battery		
					CT983	Surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F27	Problem Identified During Non-Clinical Procedure	A070504	Premature Discharge of Battery		
10689	2021	All other reportable incidents	35852	Single-chamber implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	11/02/2021	11/02/2021
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	11/02/2021	11/02/2021
							F12	Serious Injury/ Illness/ Impairment	A070504	Premature Discharge of Battery	11/02/2021	11/02/2021
							F19	Surgical Intervention	A070504	Premature Discharge of Battery	11/02/2021	11/02/2021
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	11/02/2021	11/02/2021
10690	2021	All other reportable incidents	17882	Physiologic-monitoring defibrillation system	CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070908	Failure to Sense		
							F27	Problem Identified During Non-Clinical Procedure	A070908	Failure to Sense		
10691	2021	All other reportable incidents	17882	Physiologic-monitoring defibrillation system	CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A071301	Failure to Deliver Shock/Stimulation		
							F27	Problem Identified During Non-Clinical Procedure	A071301	Failure to Deliver Shock/Stimulation		
10692	2021	All other reportable incidents	17882	Physiologic-monitoring defibrillation system	CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A071301	Failure to Deliver Shock/Stimulation		
							F27	Problem Identified During Non-Clinical Procedure	A071301	Failure to Deliver Shock/Stimulation		
10693	2021	All other reportable incidents	17882	Physiologic-monitoring defibrillation system	CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A071301	Failure to Deliver Shock/Stimulation		
							F27	Problem Identified During Non-Clinical Procedure	A071301	Failure to Deliver Shock/Stimulation		
10694	2021	All other reportable incidents	17882	Physiologic-monitoring defibrillation system	CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A071301	Failure to Deliver Shock/Stimulation		
							F27	Problem Identified During Non-Clinical Procedure	A071301	Failure to Deliver Shock/Stimulation		
10695	2021	All other reportable incidents	17882	Physiologic-monitoring defibrillation system	CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A071301	Failure to Deliver Shock/Stimulation		
							F27	Problem Identified During Non-Clinical Procedure	A071301	Failure to Deliver Shock/Stimulation		
10696	2021	All other reportable incidents	35852	Single-chamber implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	24/02/2021	24/02/2021
							F19	Surgical Intervention	A070504	Premature Discharge of Battery	24/02/2021	24/02/2021
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	24/02/2021	24/02/2021
10697	2021	All other reportable incidents	37265	Dual-chamber implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A0712	Pacing Problem		
							F26	No Health Consequences or Impact	A0712	Pacing Problem		
10698	2021	All other reportable incidents	35852	Single-chamber implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A070909	Oversensing		
							F26	No Health Consequences or Impact	A070909	Oversensing		
10699	2021	All other reportable incidents	37265	Dual-chamber implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	05/02/2021	05/02/2021
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery	05/02/2021	05/02/2021
10700	2021	All other reportable incidents	17882	Physiologic-monitoring defibrillation system	CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A071301	Failure to Deliver Shock/Stimulation		
							F27	Problem Identified During Non-Clinical Procedure	A071301	Failure to Deliver Shock/Stimulation		
10701	2021	All other reportable incidents	17882	Physiologic-monitoring defibrillation system	CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070803	Failure to Power Up		
							F27	Problem Identified During Non-Clinical Procedure	A070803	Failure to Power Up		
10702	2021	All other reportable incidents	17882	Physiologic-monitoring defibrillation system	CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070803	Failure to Power Up		
							F27	Problem Identified During Non-Clinical Procedure	A070803	Failure to Power Up		
10703	2021	All other reportable incidents	17882	Physiologic-monitoring defibrillation system	CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070803	Failure to Power Up		
							F27	Problem Identified During Non-Clinical Procedure	A070803	Failure to Power Up		
10704	2021	All other reportable incidents	63188	Left-ventricular wireless pacing system sensor/transmitter	CT141	Electrodes/Leads and associated devices	F1905	Device Revision or Replacement	A070801	Complete Loss of Power		
					CT145	Cardiovascular devices	F1905	Device Revision or Replacement	A070801	Complete Loss of Power		
					CT1577	Wireless monitoring system transmitters	F1905	Device Revision or Replacement	A070801	Complete Loss of Power		
					CT1684	Cardiac pacemakers and associated devices	F1905	Device Revision or Replacement	A070801	Complete Loss of Power		
					CT2317	Transmitters/Receivers/Transceivers	F1905	Device Revision or Replacement	A070801	Complete Loss of Power		
					CT233	Surgical	F1905	Device Revision or Replacement	A070801	Complete Loss of Power		
					CT2406	Implantable	F1905	Device Revision or Replacement	A070801	Complete Loss of Power		
					CT2407	Active implantable	F1905	Device Revision or Replacement	A070801	Complete Loss of Power		
					CT255	Transducers and associated devices	F1905	Device Revision or Replacement	A070801	Complete Loss of Power		
					CT2726	Physiological monitoring/measuring devices	F1905	Device Revision or Replacement	A070801	Complete Loss of Power		
					CT2727	Active implant devices/ancillaries	F1905	Device Revision or Replacement	A070801	Complete Loss of Power		
					CT2750	Implantable cardiac electrical stimulators and associated devices	F1905	Device Revision or Replacement	A070801	Complete Loss of Power		
					CT2885	Physiological monitoring/measuring system transmitters/receivers/transceivers	F1905	Device Revision or Replacement	A070801	Complete Loss of Power		
					CT2985	Active	F1905	Device Revision or Replacement	A070801	Complete Loss of Power		
					CT3001	Implantable metal	F1905	Device Revision or Replacement	A070801	Complete Loss of Power		
					CT3002	Implantable synthetic polymer	F1905	Device Revision or Replacement	A070801	Complete Loss of Power		
					CT3006	Implantable inorganic material	F1905	Device Revision or Replacement	A070801	Complete Loss of Power		
					CT314	Electrically powered	F1905	Device Revision or Replacement	A070801	Complete Loss of Power		
					CT321	Long-term surgical invasive	F1905	Device Revision or Replacement	A070801	Complete Loss of Power		
					CT384	Sensors and associated devices	F1905	Device Revision or Replacement	A070801	Complete Loss of Power		
					CT449	Generators and associated devices	F1905	Device Revision or Replacement	A070801	Complete Loss of Power		
					CT981	Single use	F1905	Device Revision or Replacement	A070801	Complete Loss of Power		
					CT983	Surgical invasive	F1905	Device Revision or Replacement	A070801	Complete Loss of Power		
					CT994	Sensors	F1905	Device Revision or Replacement	A070801	Complete Loss of Power		
10705	2021	All other reportable incidents	37265	Dual-chamber implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem	08/02/2021	08/02/2021
									A070504	Premature Discharge of Battery	08/02/2021	08/02/2021
							F08	Hospitalization or Prolonged Hospitalization	A0705	Battery Problem	08/02/2021	08/02/2021
									A070504	Premature Discharge of Battery	08/02/2021	08/02/2021
							F12	Serious Injury/ Illness/ Impairment	A0705	Battery Problem	08/02/2021	08/02/2021
									A070504	Premature Discharge of Battery	08/02/2021	08/02/2021

							F1903	Device Explantation	A0705	Battery Problem	08/02/2021	08/02/2021
									A070504	Premature Discharge of Battery	08/02/2021	08/02/2021
							F1905	Device Revision or Replacement	A0705	Battery Problem	08/02/2021	08/02/2021
									A070504	Premature Discharge of Battery	08/02/2021	08/02/2021
10706	2021	All other reportable incidents	35852	Single-chamber implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E2311	Discomfort	A070909	Oversensing		
							E2403	No Clinical Signs, Symptoms or Conditions	A070909	Oversensing		
							F26	No Health Consequences or Impact	A070909	Oversensing		
10707	2021	All other reportable incidents	47270	Cardiac resynchronization therapy implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A0712	Pacing Problem		
							F26	No Health Consequences or Impact	A0712	Pacing Problem		
10708	2021	All other reportable incidents	35852	Single-chamber implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	17/02/2021	17/02/2021
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	17/02/2021	17/02/2021
							F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A070504	Premature Discharge of Battery	17/02/2021	17/02/2021
							F19	Surgical Intervention	A070504	Premature Discharge of Battery	17/02/2021	17/02/2021
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	17/02/2021	17/02/2021
10709	2021	All other reportable incidents	35852	Single-chamber implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A0709	Device Sensing Problem		
							F26	No Health Consequences or Impact	A0709	Device Sensing Problem		
10710	2021	All other reportable incidents	47263	Cardiac resynchronization therapy implantable pacemaker	CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A0908	Incorrect, Inadequate or Imprecise Result or Readings		
							F26	No Health Consequences or Impact	A0908	Incorrect, Inadequate or Imprecise Result or Readings		
					CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A0908	Incorrect, Inadequate or Imprecise Result or Readings		
							F26	No Health Consequences or Impact	A0908	Incorrect, Inadequate or Imprecise Result or Readings		
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A0908	Incorrect, Inadequate or Imprecise Result or Readings		
							F26	No Health Consequences or Impact	A0908	Incorrect, Inadequate or Imprecise Result or Readings		
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A0908	Incorrect, Inadequate or Imprecise Result or Readings		
							F26	No Health Consequences or Impact	A0908	Incorrect, Inadequate or Imprecise Result or Readings		
10711	2021	All other reportable incidents	47265	Dual-chamber implantable pacemaker, rate-responsive	CT141	Electrodes/Leads and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	26/05/2021	26/05/2021
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	26/05/2021	26/05/2021
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	26/05/2021	26/05/2021
					CT145	Cardiovascular devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	26/05/2021	26/05/2021
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	26/05/2021	26/05/2021
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	26/05/2021	26/05/2021
					CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	26/05/2021	26/05/2021
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	26/05/2021	26/05/2021
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	26/05/2021	26/05/2021
					CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	26/05/2021	26/05/2021
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	26/05/2021	26/05/2021
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	26/05/2021	26/05/2021
					CT233	Surgical	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	26/05/2021	26/05/2021
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	26/05/2021	26/05/2021
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	26/05/2021	26/05/2021
					CT2406	Implantable	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	26/05/2021	26/05/2021
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	26/05/2021	26/05/2021
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	26/05/2021	26/05/2021
					CT2407	Active implantable	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	26/05/2021	26/05/2021
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	26/05/2021	26/05/2021
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	26/05/2021	26/05/2021
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	26/05/2021	26/05/2021
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	26/05/2021	26/05/2021
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	26/05/2021	26/05/2021
					CT2727	Active implant devices/ancillaries	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	26/05/2021	26/05/2021
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	26/05/2021	26/05/2021
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	26/05/2021	26/05/2021
					CT2750	Implantable cardiac electrical stimulators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	26/05/2021	26/05/2021
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	26/05/2021	26/05/2021
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	26/05/2021	26/05/2021
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	26/05/2021	26/05/2021
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	26/05/2021	26/05/2021
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	26/05/2021	26/05/2021
					CT2985	Active	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	26/05/2021	26/05/2021
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	26/05/2021	26/05/2021
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	26/05/2021	26/05/2021
					CT3001	Implantable metal	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	26/05/2021	26/05/2021
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	26/05/2021	26/05/2021
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	26/05/2021	26/05/2021
					CT3002	Implantable synthetic polymer	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	26/05/2021	26/05/2021
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	26/05/2021	26/05/2021
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	26/05/2021	26/05/2021
					CT3006	Implantable inorganic material	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	26/05/2021	26/05/2021
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	26/05/2021	26/05/2021
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	26/05/2021	26/05/2021
					CT314	Electrically powered	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	26/05/2021	26/05/2021
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	26/05/2021	26/05/2021
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	26/05/2021	26/05/2021
					CT321	Long-term surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	26/05/2021	26/05/2021
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	26/05/2021	26/05/2021
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	26/05/2021	26/05/2021
					CT449	Generators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	26/05/2021	26/05/2021
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	26/05/2021	26/05/2021
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	26/05/2021	26/05/2021
					CT981	Single use	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	26/05/2021	26/05/2021
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	26/05/2021	26/05/2021
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	26/05/2021	26/05/2021
					CT983	Surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	26/05/2021	26/05/2021

							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	26/05/2021	26/05/2021
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	26/05/2021	26/05/2021
10712	2021	All other reportable incidents	35852	Single-chamber implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A070909	Oversensing		
							F26	No Health Consequences or Impact	A070909	Oversensing		
10713	2021	All other reportable incidents	17882	Physiologic-monitoring defibrillation system	CT1684	Cardiac pacemakers and associated devices			A0708	Power Problem		
							E2403	No Clinical Signs, Symptoms or Conditions	A0708	Power Problem		
							F27	Problem Identified During Non-Clinical Procedure	A0708	Power Problem		
10714	2021	All other reportable incidents	37265	Dual-chamber implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A070910	Undersensing	03/03/2021	03/03/2021
									A071301	Failure to Deliver Shock/Stimulation	03/03/2021	03/03/2021
									A071302	Inappropriate/Inadequate Shock/Stimulation	03/03/2021	03/03/2021
							F12	Serious Injury/ Illness/ Impairment	A070910	Undersensing	03/03/2021	03/03/2021
									A071301	Failure to Deliver Shock/Stimulation	03/03/2021	03/03/2021
									A071302	Inappropriate/Inadequate Shock/Stimulation	03/03/2021	03/03/2021
10715	2021	All other reportable incidents	36241	Pacemaker/defibrillator lead	CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070101	Failure to Capture		
									A070909	Oversensing		
									A090801	Signal Artifact/Noise		
							F26	No Health Consequences or Impact	A070101	Failure to Capture		
									A070909	Oversensing		
									A090801	Signal Artifact/Noise		
10716	2021	All other reportable incidents	36241	Pacemaker/defibrillator lead	CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070909	Oversensing		
									A090801	Signal Artifact/Noise		
							F26	No Health Consequences or Impact	A070909	Oversensing		
									A090801	Signal Artifact/Noise		
10717	2021	All other reportable incidents	36241	Pacemaker/defibrillator lead	CT1684	Cardiac pacemakers and associated devices	E2104	Electric Shock	A070909	Oversensing		
									A071302	Inappropriate/Inadequate Shock/Stimulation		
									A072201	High Impedance		
							F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A070909	Oversensing		
									A071302	Inappropriate/Inadequate Shock/Stimulation		
									A072201	High Impedance		
10718	2021	All other reportable incidents	35852	Single-chamber implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	13/03/2021	13/03/2021
							F19	Surgical Intervention	A070504	Premature Discharge of Battery	13/03/2021	13/03/2021
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	13/03/2021	13/03/2021
10719	2021	All other reportable incidents	17882	Physiologic-monitoring defibrillation system	CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070803	Failure to Power Up		
							F27	Problem Identified During Non-Clinical Procedure	A070803	Failure to Power Up		
10720	2021	All other reportable incidents	47265	Dual-chamber implantable pacemaker, rate-responsive	CT141	Electrodes/Leads and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	23/02/2021	23/02/2021
							F19	Surgical Intervention	A070504	Premature Discharge of Battery	23/02/2021	23/02/2021
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	23/02/2021	23/02/2021
					CT145	Cardiovascular devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	23/02/2021	23/02/2021
							F19	Surgical Intervention	A070504	Premature Discharge of Battery	23/02/2021	23/02/2021
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	23/02/2021	23/02/2021
					CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	23/02/2021	23/02/2021
							F19	Surgical Intervention	A070504	Premature Discharge of Battery	23/02/2021	23/02/2021
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	23/02/2021	23/02/2021
					CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	23/02/2021	23/02/2021
							F19	Surgical Intervention	A070504	Premature Discharge of Battery	23/02/2021	23/02/2021
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	23/02/2021	23/02/2021
					CT233	Surgical	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	23/02/2021	23/02/2021
							F19	Surgical Intervention	A070504	Premature Discharge of Battery	23/02/2021	23/02/2021
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	23/02/2021	23/02/2021
					CT2406	Implantable	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	23/02/2021	23/02/2021
							F19	Surgical Intervention	A070504	Premature Discharge of Battery	23/02/2021	23/02/2021
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	23/02/2021	23/02/2021
					CT2407	Active implantable	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	23/02/2021	23/02/2021
							F19	Surgical Intervention	A070504	Premature Discharge of Battery	23/02/2021	23/02/2021
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	23/02/2021	23/02/2021
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	23/02/2021	23/02/2021
							F19	Surgical Intervention	A070504	Premature Discharge of Battery	23/02/2021	23/02/2021
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	23/02/2021	23/02/2021
					CT2727	Active implant devices/ancillaries	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	23/02/2021	23/02/2021
							F19	Surgical Intervention	A070504	Premature Discharge of Battery	23/02/2021	23/02/2021
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	23/02/2021	23/02/2021
					CT2750	Implantable cardiac electrical stimulators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	23/02/2021	23/02/2021
							F19	Surgical Intervention	A070504	Premature Discharge of Battery	23/02/2021	23/02/2021
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	23/02/2021	23/02/2021
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	23/02/2021	23/02/2021
							F19	Surgical Intervention	A070504	Premature Discharge of Battery	23/02/2021	23/02/2021
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	23/02/2021	23/02/2021
					CT2985	Active	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	23/02/2021	23/02/2021
							F19	Surgical Intervention	A070504	Premature Discharge of Battery	23/02/2021	23/02/2021
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	23/02/2021	23/02/2021
					CT3001	Implantable metal	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	23/02/2021	23/02/2021
							F19	Surgical Intervention	A070504	Premature Discharge of Battery	23/02/2021	23/02/2021
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	23/02/2021	23/02/2021
					CT3002	Implantable synthetic polymer	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	23/02/2021	23/02/2021
							F19	Surgical Intervention	A070504	Premature Discharge of Battery	23/02/2021	23/02/2021
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	23/02/2021	23/02/2021
					CT3006	Implantable inorganic material	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	23/02/2021	23/02/2021
							F19	Surgical Intervention	A070504	Premature Discharge of Battery	23/02/2021	23/02/2021
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	23/02/2021	23/02/2021
					CT314	Electrically powered	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	23/02/2021	23/02/2021
							F19	Surgical Intervention	A070504	Premature Discharge of Battery	23/02/2021	23/02/2021
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	23/02/2021	23/02/2021
					CT321	Long-term surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	23/02/2021	23/02/2021
							F19	Surgical Intervention	A070504	Premature Discharge of Battery	23/02/2021	23/02/2021
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	23/02/2021	23/02/2021
					CT449	Generators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	23/02/2021	23/02/2021
							F19	Surgical Intervention	A070504	Premature Discharge of Battery	23/02/2021	23/02/2021
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	23/02/2021	23/02/2021

					CT981	Single use	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	23/02/2021	23/02/2021
							F19	Surgical Intervention	A070504	Premature Discharge of Battery	23/02/2021	23/02/2021
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	23/02/2021	23/02/2021
					CT983	Surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	23/02/2021	23/02/2021
							F19	Surgical Intervention	A070504	Premature Discharge of Battery	23/02/2021	23/02/2021
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	23/02/2021	23/02/2021
10721	2021	All other reportable incidents	36241	Pacemaker/defibrillator lead	CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A0203	Defective Device	19/02/2021	19/02/2021
									A23	Use of Device Problem	19/02/2021	19/02/2021
							F19	Surgical Intervention	A0203	Defective Device	19/02/2021	19/02/2021
									A23	Use of Device Problem	19/02/2021	19/02/2021
10722	2021	All other reportable incidents	47263	Cardiac resynchronization therapy implantable pacemaker	CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A0908	Incorrect, Inadequate or Imprecise Result or Readings	04/02/2021	04/02/2021
							F08	Hospitalization or Prolonged Hospitalization	A0908	Incorrect, Inadequate or Imprecise Result or Readings	04/02/2021	04/02/2021
							F1903	Device Explantation	A0908	Incorrect, Inadequate or Imprecise Result or Readings	04/02/2021	04/02/2021
							F1905	Device Revision or Replacement	A0908	Incorrect, Inadequate or Imprecise Result or Readings	04/02/2021	04/02/2021
					CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A0908	Incorrect, Inadequate or Imprecise Result or Readings	04/02/2021	04/02/2021
							F08	Hospitalization or Prolonged Hospitalization	A0908	Incorrect, Inadequate or Imprecise Result or Readings	04/02/2021	04/02/2021
							F1903	Device Explantation	A0908	Incorrect, Inadequate or Imprecise Result or Readings	04/02/2021	04/02/2021
							F1905	Device Revision or Replacement	A0908	Incorrect, Inadequate or Imprecise Result or Readings	04/02/2021	04/02/2021
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A0908	Incorrect, Inadequate or Imprecise Result or Readings	04/02/2021	04/02/2021
							F08	Hospitalization or Prolonged Hospitalization	A0908	Incorrect, Inadequate or Imprecise Result or Readings	04/02/2021	04/02/2021
							F1903	Device Explantation	A0908	Incorrect, Inadequate or Imprecise Result or Readings	04/02/2021	04/02/2021
							F1905	Device Revision or Replacement	A0908	Incorrect, Inadequate or Imprecise Result or Readings	04/02/2021	04/02/2021
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A0908	Incorrect, Inadequate or Imprecise Result or Readings	04/02/2021	04/02/2021
							F08	Hospitalization or Prolonged Hospitalization	A0908	Incorrect, Inadequate or Imprecise Result or Readings	04/02/2021	04/02/2021
							F1903	Device Explantation	A0908	Incorrect, Inadequate or Imprecise Result or Readings	04/02/2021	04/02/2021
							F1905	Device Revision or Replacement	A0908	Incorrect, Inadequate or Imprecise Result or Readings	04/02/2021	04/02/2021
10723	2021	All other reportable incidents	36241	Pacemaker/defibrillator lead	CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A072201	High Impedance		
							F26	No Health Consequences or Impact	A072201	High Impedance		
10724	2021	All other reportable incidents	17882	Physiologic-monitoring defibrillation system	CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A07	Electrical /Electronic Property Problem		
							F26	No Health Consequences or Impact	A07	Electrical /Electronic Property Problem		
10725	2021	All other reportable incidents	35852	Single-chamber implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem	12/03/2021	12/03/2021
									A070504	Premature Discharge of Battery	12/03/2021	12/03/2021
							F19	Surgical Intervention	A0705	Battery Problem	12/03/2021	12/03/2021
									A070504	Premature Discharge of Battery	12/03/2021	12/03/2021
10726	2021	All other reportable incidents	36241	Pacemaker/defibrillator lead	CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A072201	High Impedance		
							F26	No Health Consequences or Impact	A072201	High Impedance		
10727	2021	All other reportable incidents	47265	Dual-chamber implantable pacemaker, rate-responsive	CT141	Electrodes/Leads and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem		
							F26	No Health Consequences or Impact	A0705	Battery Problem		
					CT145	Cardiovascular devices	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem		
							F26	No Health Consequences or Impact	A0705	Battery Problem		
					CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem		
							F26	No Health Consequences or Impact	A0705	Battery Problem		
					CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem		
							F26	No Health Consequences or Impact	A0705	Battery Problem		
					CT233	Surgical	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem		
							F26	No Health Consequences or Impact	A0705	Battery Problem		
					CT2406	Implantable	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem		
							F26	No Health Consequences or Impact	A0705	Battery Problem		
					CT2407	Active implantable	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem		
							F26	No Health Consequences or Impact	A0705	Battery Problem		
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem		
							F26	No Health Consequences or Impact	A0705	Battery Problem		
					CT2727	Active implant devices/ancillaries	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem		
							F26	No Health Consequences or Impact	A0705	Battery Problem		
					CT2750	Implantable cardiac electrical stimulators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem		
							F26	No Health Consequences or Impact	A0705	Battery Problem		
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem		
							F26	No Health Consequences or Impact	A0705	Battery Problem		
					CT2985	Active	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem		
							F26	No Health Consequences or Impact	A0705	Battery Problem		
					CT3001	Implantable metal	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem		
							F26	No Health Consequences or Impact	A0705	Battery Problem		
					CT3002	Implantable synthetic polymer	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem		
							F26	No Health Consequences or Impact	A0705	Battery Problem		
					CT3006	Implantable inorganic material	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem		
							F26	No Health Consequences or Impact	A0705	Battery Problem		
					CT314	Electrically powered	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem		
							F26	No Health Consequences or Impact	A0705	Battery Problem		
					CT321	Long-term surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem		
							F26	No Health Consequences or Impact	A0705	Battery Problem		
					CT449	Generators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem		
							F26	No Health Consequences or Impact	A0705	Battery Problem		
					CT981	Single use	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem		
							F26	No Health Consequences or Impact	A0705	Battery Problem		
					CT983	Surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem		
							F26	No Health Consequences or Impact	A0705	Battery Problem		

10728	2021	All other reportable incidents	47265	Dual-chamber implantable pacemaker, rate-responsive	CT141	Electrodes/Leads and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT145	Cardiovascular devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT233	Surgical	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT2406	Implantable	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT2407	Active implantable	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT2727	Active implant devices/ancillaries	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT2750	Implantable cardiac electrical stimulators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT2985	Active	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT3001	Implantable metal	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT3002	Implantable synthetic polymer	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT3006	Implantable inorganic material	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT314	Electrically powered	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT321	Long-term surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT449	Generators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT981	Single use	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT983	Surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
10729	2021	All other reportable incidents	47267	Single-chamber implantable pacemaker, rate-responsive	CT141	Electrodes/Leads and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	17/02/2021	17/02/2021
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	17/02/2021	17/02/2021
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	17/02/2021	17/02/2021
					CT145	Cardiovascular devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	17/02/2021	17/02/2021
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	17/02/2021	17/02/2021
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	17/02/2021	17/02/2021
					CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	17/02/2021	17/02/2021
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	17/02/2021	17/02/2021
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	17/02/2021	17/02/2021
					CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	17/02/2021	17/02/2021
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	17/02/2021	17/02/2021
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	17/02/2021	17/02/2021
					CT233	Surgical	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	17/02/2021	17/02/2021
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	17/02/2021	17/02/2021
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	17/02/2021	17/02/2021
					CT2406	Implantable	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	17/02/2021	17/02/2021
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	17/02/2021	17/02/2021
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	17/02/2021	17/02/2021
					CT2407	Active implantable	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	17/02/2021	17/02/2021
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	17/02/2021	17/02/2021
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	17/02/2021	17/02/2021
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	17/02/2021	17/02/2021
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	17/02/2021	17/02/2021
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	17/02/2021	17/02/2021
					CT2727	Active implant devices/ancillaries	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	17/02/2021	17/02/2021
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	17/02/2021	17/02/2021
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	17/02/2021	17/02/2021
					CT2750	Implantable cardiac electrical stimulators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	17/02/2021	17/02/2021
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	17/02/2021	17/02/2021
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	17/02/2021	17/02/2021
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	17/02/2021	17/02/2021
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	17/02/2021	17/02/2021
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	17/02/2021	17/02/2021
					CT2985	Active	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	17/02/2021	17/02/2021
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	17/02/2021	17/02/2021
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	17/02/2021	17/02/2021
					CT3001	Implantable metal	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	17/02/2021	17/02/2021
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	17/02/2021	17/02/2021
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	17/02/2021	17/02/2021
					CT3002	Implantable synthetic polymer	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	17/02/2021	17/02/2021
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	17/02/2021	17/02/2021
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	17/02/2021	17/02/2021
					CT3006	Implantable inorganic material	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	17/02/2021	17/02/2021
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	17/02/2021	17/02/2021
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	17/02/2021	17/02/2021
					CT314	Electrically powered	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	17/02/2021	17/02/2021
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	17/02/2021	17/02/2021
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	17/02/2021	17/02/2021

					CT321	Long-term surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	17/02/2021	17/02/2021
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	17/02/2021	17/02/2021
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	17/02/2021	17/02/2021
					CT449	Generators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	17/02/2021	17/02/2021
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	17/02/2021	17/02/2021
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	17/02/2021	17/02/2021
					CT981	Single use	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	17/02/2021	17/02/2021
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	17/02/2021	17/02/2021
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	17/02/2021	17/02/2021
					CT983	Surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	17/02/2021	17/02/2021
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	17/02/2021	17/02/2021
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	17/02/2021	17/02/2021
10730	2021	All other reportable incidents	47267	Single-chamber implantable pacemaker, rate-responsive	CT141	Electrodes/Leads and associated devices	E010701	Confusion/ Disorientation	A0712	Pacing Problem	26/02/2021	26/02/2021
							E011903	Syncope/Fainting	A0712	Pacing Problem	26/02/2021	26/02/2021
							E060101	Asystole	A0712	Pacing Problem	26/02/2021	26/02/2021
							E0602	Cardiac Arrest	A0712	Pacing Problem	26/02/2021	26/02/2021
							F08	Hospitalization or Prolonged Hospitalization	A0712	Pacing Problem	26/02/2021	26/02/2021
							F1203	Life Threatening Illness or Injury	A0712	Pacing Problem	26/02/2021	26/02/2021
							F1905	Device Revision or Replacement	A0712	Pacing Problem	26/02/2021	26/02/2021
					CT145	Cardiovascular devices	E010701	Confusion/ Disorientation	A0712	Pacing Problem	26/02/2021	26/02/2021
							E011903	Syncope/Fainting	A0712	Pacing Problem	26/02/2021	26/02/2021
							E060101	Asystole	A0712	Pacing Problem	26/02/2021	26/02/2021
							E0602	Cardiac Arrest	A0712	Pacing Problem	26/02/2021	26/02/2021
							F08	Hospitalization or Prolonged Hospitalization	A0712	Pacing Problem	26/02/2021	26/02/2021
							F1203	Life Threatening Illness or Injury	A0712	Pacing Problem	26/02/2021	26/02/2021
							F1905	Device Revision or Replacement	A0712	Pacing Problem	26/02/2021	26/02/2021
					CT1684	Cardiac pacemakers and associated devices	E010701	Confusion/ Disorientation	A0712	Pacing Problem	26/02/2021	26/02/2021
							E011903	Syncope/Fainting	A0712	Pacing Problem	26/02/2021	26/02/2021
							E060101	Asystole	A0712	Pacing Problem	26/02/2021	26/02/2021
							E0602	Cardiac Arrest	A0712	Pacing Problem	26/02/2021	26/02/2021
							F08	Hospitalization or Prolonged Hospitalization	A0712	Pacing Problem	26/02/2021	26/02/2021
							F1203	Life Threatening Illness or Injury	A0712	Pacing Problem	26/02/2021	26/02/2021
							F1905	Device Revision or Replacement	A0712	Pacing Problem	26/02/2021	26/02/2021
					CT1685	Cardiac pacemakers	E010701	Confusion/ Disorientation	A0712	Pacing Problem	26/02/2021	26/02/2021
							E011903	Syncope/Fainting	A0712	Pacing Problem	26/02/2021	26/02/2021
							E060101	Asystole	A0712	Pacing Problem	26/02/2021	26/02/2021
							E0602	Cardiac Arrest	A0712	Pacing Problem	26/02/2021	26/02/2021
							F08	Hospitalization or Prolonged Hospitalization	A0712	Pacing Problem	26/02/2021	26/02/2021
							F1203	Life Threatening Illness or Injury	A0712	Pacing Problem	26/02/2021	26/02/2021
							F1905	Device Revision or Replacement	A0712	Pacing Problem	26/02/2021	26/02/2021
					CT233	Surgical	E010701	Confusion/ Disorientation	A0712	Pacing Problem	26/02/2021	26/02/2021
							E011903	Syncope/Fainting	A0712	Pacing Problem	26/02/2021	26/02/2021
							E060101	Asystole	A0712	Pacing Problem	26/02/2021	26/02/2021
							E0602	Cardiac Arrest	A0712	Pacing Problem	26/02/2021	26/02/2021
							F08	Hospitalization or Prolonged Hospitalization	A0712	Pacing Problem	26/02/2021	26/02/2021
							F1203	Life Threatening Illness or Injury	A0712	Pacing Problem	26/02/2021	26/02/2021
							F1905	Device Revision or Replacement	A0712	Pacing Problem	26/02/2021	26/02/2021
					CT2406	Implantable	E010701	Confusion/ Disorientation	A0712	Pacing Problem	26/02/2021	26/02/2021
							E011903	Syncope/Fainting	A0712	Pacing Problem	26/02/2021	26/02/2021
							E060101	Asystole	A0712	Pacing Problem	26/02/2021	26/02/2021
							E0602	Cardiac Arrest	A0712	Pacing Problem	26/02/2021	26/02/2021
							F08	Hospitalization or Prolonged Hospitalization	A0712	Pacing Problem	26/02/2021	26/02/2021
							F1203	Life Threatening Illness or Injury	A0712	Pacing Problem	26/02/2021	26/02/2021
							F1905	Device Revision or Replacement	A0712	Pacing Problem	26/02/2021	26/02/2021
					CT2407	Active implantable	E010701	Confusion/ Disorientation	A0712	Pacing Problem	26/02/2021	26/02/2021
							E011903	Syncope/Fainting	A0712	Pacing Problem	26/02/2021	26/02/2021
							E060101	Asystole	A0712	Pacing Problem	26/02/2021	26/02/2021
							E0602	Cardiac Arrest	A0712	Pacing Problem	26/02/2021	26/02/2021
							F08	Hospitalization or Prolonged Hospitalization	A0712	Pacing Problem	26/02/2021	26/02/2021
							F1203	Life Threatening Illness or Injury	A0712	Pacing Problem	26/02/2021	26/02/2021
							F1905	Device Revision or Replacement	A0712	Pacing Problem	26/02/2021	26/02/2021
					CT257	Implantable cardiac pacemakers	E010701	Confusion/ Disorientation	A0712	Pacing Problem	26/02/2021	26/02/2021
							E011903	Syncope/Fainting	A0712	Pacing Problem	26/02/2021	26/02/2021
							E060101	Asystole	A0712	Pacing Problem	26/02/2021	26/02/2021
							E0602	Cardiac Arrest	A0712	Pacing Problem	26/02/2021	26/02/2021
							F08	Hospitalization or Prolonged Hospitalization	A0712	Pacing Problem	26/02/2021	26/02/2021
							F1203	Life Threatening Illness or Injury	A0712	Pacing Problem	26/02/2021	26/02/2021
							F1905	Device Revision or Replacement	A0712	Pacing Problem	26/02/2021	26/02/2021
					CT2727	Active implant devices/ancillaries	E010701	Confusion/ Disorientation	A0712	Pacing Problem	26/02/2021	26/02/2021
							E011903	Syncope/Fainting	A0712	Pacing Problem	26/02/2021	26/02/2021
							E060101	Asystole	A0712	Pacing Problem	26/02/2021	26/02/2021
							E0602	Cardiac Arrest	A0712	Pacing Problem	26/02/2021	26/02/2021
							F08	Hospitalization or Prolonged Hospitalization	A0712	Pacing Problem	26/02/2021	26/02/2021
							F1203	Life Threatening Illness or Injury	A0712	Pacing Problem	26/02/2021	26/02/2021
							F1905	Device Revision or Replacement	A0712	Pacing Problem	26/02/2021	26/02/2021
					CT2750	Implantable cardiac electrical stimulators and associated devices	E010701	Confusion/ Disorientation	A0712	Pacing Problem	26/02/2021	26/02/2021
							E011903	Syncope/Fainting	A0712	Pacing Problem	26/02/2021	26/02/2021
							E060101	Asystole	A0712	Pacing Problem	26/02/2021	26/02/2021
							E0602	Cardiac Arrest	A0712	Pacing Problem	26/02/2021	26/02/2021
							F08	Hospitalization or Prolonged Hospitalization	A0712	Pacing Problem	26/02/2021	26/02/2021
							F1203	Life Threatening Illness or Injury	A0712	Pacing Problem	26/02/2021	26/02/2021
							F1905	Device Revision or Replacement	A0712	Pacing Problem	26/02/2021	26/02/2021
					CT2866	Implantable cardiac electrical stimulators	E010701	Confusion/ Disorientation	A0712	Pacing Problem	26/02/2021	26/02/2021
							E011903	Syncope/Fainting	A0712	Pacing Problem	26/02/2021	26/02/2021
							E060101	Asystole	A0712	Pacing Problem	26/02/2021	26/02/2021
							E0602	Cardiac Arrest	A0712	Pacing Problem	26/02/2021	26/02/2021
							F08	Hospitalization or Prolonged Hospitalization	A0712	Pacing Problem	26/02/2021	26/02/2021
							F1203	Life Threatening Illness or Injury	A0712	Pacing Problem	26/02/2021	26/02/2021

							F1905	Device Revision or Replacement	A0712	Pacing Problem	26/02/2021	26/02/2021
					CT2985	Active	E010701	Confusion/ Disorientation	A0712	Pacing Problem	26/02/2021	26/02/2021
							E011903	Syncope/Fainting	A0712	Pacing Problem	26/02/2021	26/02/2021
							E060101	Asystole	A0712	Pacing Problem	26/02/2021	26/02/2021
							E0602	Cardiac Arrest	A0712	Pacing Problem	26/02/2021	26/02/2021
							F08	Hospitalization or Prolonged Hospitalization	A0712	Pacing Problem	26/02/2021	26/02/2021
							F1203	Life Threatening Illness or Injury	A0712	Pacing Problem	26/02/2021	26/02/2021
							F1905	Device Revision or Replacement	A0712	Pacing Problem	26/02/2021	26/02/2021
					CT3001	Implantable metal	E010701	Confusion/ Disorientation	A0712	Pacing Problem	26/02/2021	26/02/2021
							E011903	Syncope/Fainting	A0712	Pacing Problem	26/02/2021	26/02/2021
							E060101	Asystole	A0712	Pacing Problem	26/02/2021	26/02/2021
							E0602	Cardiac Arrest	A0712	Pacing Problem	26/02/2021	26/02/2021
							F08	Hospitalization or Prolonged Hospitalization	A0712	Pacing Problem	26/02/2021	26/02/2021
							F1203	Life Threatening Illness or Injury	A0712	Pacing Problem	26/02/2021	26/02/2021
							F1905	Device Revision or Replacement	A0712	Pacing Problem	26/02/2021	26/02/2021
					CT3002	Implantable synthetic polymer	E010701	Confusion/ Disorientation	A0712	Pacing Problem	26/02/2021	26/02/2021
							E011903	Syncope/Fainting	A0712	Pacing Problem	26/02/2021	26/02/2021
							E060101	Asystole	A0712	Pacing Problem	26/02/2021	26/02/2021
							E0602	Cardiac Arrest	A0712	Pacing Problem	26/02/2021	26/02/2021
							F08	Hospitalization or Prolonged Hospitalization	A0712	Pacing Problem	26/02/2021	26/02/2021
							F1203	Life Threatening Illness or Injury	A0712	Pacing Problem	26/02/2021	26/02/2021
							F1905	Device Revision or Replacement	A0712	Pacing Problem	26/02/2021	26/02/2021
					CT3006	Implantable inorganic material	E010701	Confusion/ Disorientation	A0712	Pacing Problem	26/02/2021	26/02/2021
							E011903	Syncope/Fainting	A0712	Pacing Problem	26/02/2021	26/02/2021
							E060101	Asystole	A0712	Pacing Problem	26/02/2021	26/02/2021
							E0602	Cardiac Arrest	A0712	Pacing Problem	26/02/2021	26/02/2021
							F08	Hospitalization or Prolonged Hospitalization	A0712	Pacing Problem	26/02/2021	26/02/2021
							F1203	Life Threatening Illness or Injury	A0712	Pacing Problem	26/02/2021	26/02/2021
							F1905	Device Revision or Replacement	A0712	Pacing Problem	26/02/2021	26/02/2021
					CT314	Electrically powered	E010701	Confusion/ Disorientation	A0712	Pacing Problem	26/02/2021	26/02/2021
							E011903	Syncope/Fainting	A0712	Pacing Problem	26/02/2021	26/02/2021
							E060101	Asystole	A0712	Pacing Problem	26/02/2021	26/02/2021
							E0602	Cardiac Arrest	A0712	Pacing Problem	26/02/2021	26/02/2021
							F08	Hospitalization or Prolonged Hospitalization	A0712	Pacing Problem	26/02/2021	26/02/2021
							F1203	Life Threatening Illness or Injury	A0712	Pacing Problem	26/02/2021	26/02/2021
							F1905	Device Revision or Replacement	A0712	Pacing Problem	26/02/2021	26/02/2021
					CT321	Long-term surgical invasive	E010701	Confusion/ Disorientation	A0712	Pacing Problem	26/02/2021	26/02/2021
							E011903	Syncope/Fainting	A0712	Pacing Problem	26/02/2021	26/02/2021
							E060101	Asystole	A0712	Pacing Problem	26/02/2021	26/02/2021
							E0602	Cardiac Arrest	A0712	Pacing Problem	26/02/2021	26/02/2021
							F08	Hospitalization or Prolonged Hospitalization	A0712	Pacing Problem	26/02/2021	26/02/2021
							F1203	Life Threatening Illness or Injury	A0712	Pacing Problem	26/02/2021	26/02/2021
							F1905	Device Revision or Replacement	A0712	Pacing Problem	26/02/2021	26/02/2021
					CT449	Generators and associated devices	E010701	Confusion/ Disorientation	A0712	Pacing Problem	26/02/2021	26/02/2021
							E011903	Syncope/Fainting	A0712	Pacing Problem	26/02/2021	26/02/2021
							E060101	Asystole	A0712	Pacing Problem	26/02/2021	26/02/2021
							E0602	Cardiac Arrest	A0712	Pacing Problem	26/02/2021	26/02/2021
							F08	Hospitalization or Prolonged Hospitalization	A0712	Pacing Problem	26/02/2021	26/02/2021
							F1203	Life Threatening Illness or Injury	A0712	Pacing Problem	26/02/2021	26/02/2021
							F1905	Device Revision or Replacement	A0712	Pacing Problem	26/02/2021	26/02/2021
					CT981	Single use	E010701	Confusion/ Disorientation	A0712	Pacing Problem	26/02/2021	26/02/2021
							E011903	Syncope/Fainting	A0712	Pacing Problem	26/02/2021	26/02/2021
							E060101	Asystole	A0712	Pacing Problem	26/02/2021	26/02/2021
							E0602	Cardiac Arrest	A0712	Pacing Problem	26/02/2021	26/02/2021
							F08	Hospitalization or Prolonged Hospitalization	A0712	Pacing Problem	26/02/2021	26/02/2021
							F1203	Life Threatening Illness or Injury	A0712	Pacing Problem	26/02/2021	26/02/2021
							F1905	Device Revision or Replacement	A0712	Pacing Problem	26/02/2021	26/02/2021
					CT983	Surgical invasive	E010701	Confusion/ Disorientation	A0712	Pacing Problem	26/02/2021	26/02/2021
							E011903	Syncope/Fainting	A0712	Pacing Problem	26/02/2021	26/02/2021
							E060101	Asystole	A0712	Pacing Problem	26/02/2021	26/02/2021
							E0602	Cardiac Arrest	A0712	Pacing Problem	26/02/2021	26/02/2021
							F08	Hospitalization or Prolonged Hospitalization	A0712	Pacing Problem	26/02/2021	26/02/2021
							F1203	Life Threatening Illness or Injury	A0712	Pacing Problem	26/02/2021	26/02/2021
							F1905	Device Revision or Replacement	A0712	Pacing Problem	26/02/2021	26/02/2021
10731	2021	All other reportable incidents	35852	Single-chamber implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F15	Recognized Device or Procedural Complication	A070504	Premature Discharge of Battery		
10732	2021	All other reportable incidents	17882	Physiologic-monitoring defibrillation system	CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A0719	Unexpected Shutdown		
							F26	No Health Consequences or Impact	A0719	Unexpected Shutdown		
10733	2021	All other reportable incidents	17882	Physiologic-monitoring defibrillation system	CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070803	Failure to Power Up		
							F27	Problem Identified During Non-Clinical Procedure	A070803	Failure to Power Up		
10734	2021	Death	47267	Single-chamber implantable pacemaker, rate-responsive	CT141	Electrodes/Leads and associated devices	E2401	Insufficient Information	A24	No Apparent Device Problem		
							F02	Death	A24	No Apparent Device Problem		
					CT145	Cardiovascular devices	E2401	Insufficient Information	A24	No Apparent Device Problem		
							F02	Death	A24	No Apparent Device Problem		
					CT1684	Cardiac pacemakers and associated devices	E2401	Insufficient Information	A24	No Apparent Device Problem		
							F02	Death	A24	No Apparent Device Problem		
					CT1685	Cardiac pacemakers	E2401	Insufficient Information	A24	No Apparent Device Problem		
							F02	Death	A24	No Apparent Device Problem		
					CT233	Surgical	E2401	Insufficient Information	A24	No Apparent Device Problem		
							F02	Death	A24	No Apparent Device Problem		
					CT2406	Implantable	E2401	Insufficient Information	A24	No Apparent Device Problem		
							F02	Death	A24	No Apparent Device Problem		
					CT2407	Active implantable	E2401	Insufficient Information	A24	No Apparent Device Problem		
							F02	Death	A24	No Apparent Device Problem		
					CT257	Implantable cardiac pacemakers	E2401	Insufficient Information	A24	No Apparent Device Problem		
							F02	Death	A24	No Apparent Device Problem		
					CT2727	Active implant devices/ancillaries	E2401	Insufficient Information	A24	No Apparent Device Problem		
							F02	Death	A24	No Apparent Device Problem		

					CT2750	Implantable cardiac electrical stimulators and associated devices	E2401	Insufficient Information	A24	No Apparent Device Problem		
							F02	Death	A24	No Apparent Device Problem		
					CT2866	Implantable cardiac electrical stimulators	E2401	Insufficient Information	A24	No Apparent Device Problem		
							F02	Death	A24	No Apparent Device Problem		
					CT2985	Active	E2401	Insufficient Information	A24	No Apparent Device Problem		
							F02	Death	A24	No Apparent Device Problem		
					CT3001	Implantable metal	E2401	Insufficient Information	A24	No Apparent Device Problem		
							F02	Death	A24	No Apparent Device Problem		
					CT3002	Implantable synthetic polymer	E2401	Insufficient Information	A24	No Apparent Device Problem		
							F02	Death	A24	No Apparent Device Problem		
					CT3006	Implantable inorganic material	E2401	Insufficient Information	A24	No Apparent Device Problem		
							F02	Death	A24	No Apparent Device Problem		
					CT314	Electrically powered	E2401	Insufficient Information	A24	No Apparent Device Problem		
							F02	Death	A24	No Apparent Device Problem		
					CT321	Long-term surgical invasive	E2401	Insufficient Information	A24	No Apparent Device Problem		
							F02	Death	A24	No Apparent Device Problem		
					CT449	Generators and associated devices	E2401	Insufficient Information	A24	No Apparent Device Problem		
							F02	Death	A24	No Apparent Device Problem		
					CT981	Single use	E2401	Insufficient Information	A24	No Apparent Device Problem		
							F02	Death	A24	No Apparent Device Problem		
					CT983	Surgical invasive	E2401	Insufficient Information	A24	No Apparent Device Problem		
							F02	Death	A24	No Apparent Device Problem		
10735	2021	All other reportable incidents	35852	Single-chamber implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	11/02/2021	11/02/2021
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	11/02/2021	11/02/2021
							F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A070504	Premature Discharge of Battery	11/02/2021	11/02/2021
							F12	Serious Injury/ Illness/ Impairment	A070504	Premature Discharge of Battery	11/02/2021	11/02/2021
							F19	Surgical Intervention	A070504	Premature Discharge of Battery	11/02/2021	11/02/2021
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	11/02/2021	11/02/2021
10736	2021	All other reportable incidents	47267	Single-chamber implantable pacemaker, rate-responsive	CT141	Electrodes/Leads and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT145	Cardiovascular devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT233	Surgical	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT2406	Implantable	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT2407	Active implantable	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT2727	Active implant devices/ancillaries	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT2750	Implantable cardiac electrical stimulators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT2985	Active	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT3001	Implantable metal	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT3002	Implantable synthetic polymer	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT3006	Implantable inorganic material	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT314	Electrically powered	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT321	Long-term surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT449	Generators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT981	Single use	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT983	Surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
10737	2021	All other reportable incidents	47267	Single-chamber implantable pacemaker, rate-responsive	CT141	Electrodes/Leads and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT145	Cardiovascular devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT233	Surgical	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT2406	Implantable	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT2407	Active implantable	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT2727	Active implant devices/ancillaries	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT2750	Implantable cardiac electrical stimulators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		

					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT2985	Active	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT3001	Implantable metal	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT3002	Implantable synthetic polymer	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT3006	Implantable inorganic material	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT314	Electrically powered	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT321	Long-term surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT449	Generators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT981	Single use	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT983	Surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
10738	2021	All other reportable incidents	47265	Dual-chamber implantable pacemaker, rate-responsive	CT141	Electrodes/Leads and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A0712	Pacing Problem	03/03/2021	03/03/2021
							F1905	Device Revision or Replacement	A0712	Pacing Problem	03/03/2021	03/03/2021
							F26	No Health Consequences or Impact	A0712	Pacing Problem	03/03/2021	03/03/2021
					CT145	Cardiovascular devices	E2403	No Clinical Signs, Symptoms or Conditions	A0712	Pacing Problem	03/03/2021	03/03/2021
							F1905	Device Revision or Replacement	A0712	Pacing Problem	03/03/2021	03/03/2021
							F26	No Health Consequences or Impact	A0712	Pacing Problem	03/03/2021	03/03/2021
					CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A0712	Pacing Problem	03/03/2021	03/03/2021
							F1905	Device Revision or Replacement	A0712	Pacing Problem	03/03/2021	03/03/2021
							F26	No Health Consequences or Impact	A0712	Pacing Problem	03/03/2021	03/03/2021
					CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A0712	Pacing Problem	03/03/2021	03/03/2021
							F1905	Device Revision or Replacement	A0712	Pacing Problem	03/03/2021	03/03/2021
							F26	No Health Consequences or Impact	A0712	Pacing Problem	03/03/2021	03/03/2021
					CT233	Surgical	E2403	No Clinical Signs, Symptoms or Conditions	A0712	Pacing Problem	03/03/2021	03/03/2021
							F1905	Device Revision or Replacement	A0712	Pacing Problem	03/03/2021	03/03/2021
							F26	No Health Consequences or Impact	A0712	Pacing Problem	03/03/2021	03/03/2021
					CT2406	Implantable	E2403	No Clinical Signs, Symptoms or Conditions	A0712	Pacing Problem	03/03/2021	03/03/2021
							F1905	Device Revision or Replacement	A0712	Pacing Problem	03/03/2021	03/03/2021
							F26	No Health Consequences or Impact	A0712	Pacing Problem	03/03/2021	03/03/2021
					CT2407	Active implantable	E2403	No Clinical Signs, Symptoms or Conditions	A0712	Pacing Problem	03/03/2021	03/03/2021
							F1905	Device Revision or Replacement	A0712	Pacing Problem	03/03/2021	03/03/2021
							F26	No Health Consequences or Impact	A0712	Pacing Problem	03/03/2021	03/03/2021
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A0712	Pacing Problem	03/03/2021	03/03/2021
							F1905	Device Revision or Replacement	A0712	Pacing Problem	03/03/2021	03/03/2021
							F26	No Health Consequences or Impact	A0712	Pacing Problem	03/03/2021	03/03/2021
					CT2727	Active implant devices/ancillaries	E2403	No Clinical Signs, Symptoms or Conditions	A0712	Pacing Problem	03/03/2021	03/03/2021
							F1905	Device Revision or Replacement	A0712	Pacing Problem	03/03/2021	03/03/2021
							F26	No Health Consequences or Impact	A0712	Pacing Problem	03/03/2021	03/03/2021
					CT2750	Implantable cardiac electrical stimulators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A0712	Pacing Problem	03/03/2021	03/03/2021
							F1905	Device Revision or Replacement	A0712	Pacing Problem	03/03/2021	03/03/2021
							F26	No Health Consequences or Impact	A0712	Pacing Problem	03/03/2021	03/03/2021
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A0712	Pacing Problem	03/03/2021	03/03/2021
							F1905	Device Revision or Replacement	A0712	Pacing Problem	03/03/2021	03/03/2021
							F26	No Health Consequences or Impact	A0712	Pacing Problem	03/03/2021	03/03/2021
					CT2985	Active	E2403	No Clinical Signs, Symptoms or Conditions	A0712	Pacing Problem	03/03/2021	03/03/2021
							F1905	Device Revision or Replacement	A0712	Pacing Problem	03/03/2021	03/03/2021
							F26	No Health Consequences or Impact	A0712	Pacing Problem	03/03/2021	03/03/2021
					CT3001	Implantable metal	E2403	No Clinical Signs, Symptoms or Conditions	A0712	Pacing Problem	03/03/2021	03/03/2021
							F1905	Device Revision or Replacement	A0712	Pacing Problem	03/03/2021	03/03/2021
							F26	No Health Consequences or Impact	A0712	Pacing Problem	03/03/2021	03/03/2021
					CT3002	Implantable synthetic polymer	E2403	No Clinical Signs, Symptoms or Conditions	A0712	Pacing Problem	03/03/2021	03/03/2021
							F1905	Device Revision or Replacement	A0712	Pacing Problem	03/03/2021	03/03/2021
							F26	No Health Consequences or Impact	A0712	Pacing Problem	03/03/2021	03/03/2021
					CT3006	Implantable inorganic material	E2403	No Clinical Signs, Symptoms or Conditions	A0712	Pacing Problem	03/03/2021	03/03/2021
							F1905	Device Revision or Replacement	A0712	Pacing Problem	03/03/2021	03/03/2021
							F26	No Health Consequences or Impact	A0712	Pacing Problem	03/03/2021	03/03/2021
					CT314	Electrically powered	E2403	No Clinical Signs, Symptoms or Conditions	A0712	Pacing Problem	03/03/2021	03/03/2021
							F1905	Device Revision or Replacement	A0712	Pacing Problem	03/03/2021	03/03/2021
							F26	No Health Consequences or Impact	A0712	Pacing Problem	03/03/2021	03/03/2021
					CT321	Long-term surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A0712	Pacing Problem	03/03/2021	03/03/2021
							F1905	Device Revision or Replacement	A0712	Pacing Problem	03/03/2021	03/03/2021
							F26	No Health Consequences or Impact	A0712	Pacing Problem	03/03/2021	03/03/2021
					CT449	Generators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A0712	Pacing Problem	03/03/2021	03/03/2021
							F1905	Device Revision or Replacement	A0712	Pacing Problem	03/03/2021	03/03/2021
							F26	No Health Consequences or Impact	A0712	Pacing Problem	03/03/2021	03/03/2021
					CT981	Single use	E2403	No Clinical Signs, Symptoms or Conditions	A0712	Pacing Problem	03/03/2021	03/03/2021
							F1905	Device Revision or Replacement	A0712	Pacing Problem	03/03/2021	03/03/2021
							F26	No Health Consequences or Impact	A0712	Pacing Problem	03/03/2021	03/03/2021
					CT983	Surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A0712	Pacing Problem	03/03/2021	03/03/2021
							F1905	Device Revision or Replacement	A0712	Pacing Problem	03/03/2021	03/03/2021
							F26	No Health Consequences or Impact	A0712	Pacing Problem	03/03/2021	03/03/2021
10739	2021	All other reportable incidents	36241	Pacemaker/defibrillator lead	CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A0203	Defective Device	11/02/2021	11/02/2021
							F08	Hospitalization or Prolonged Hospitalization	A0203	Defective Device	11/02/2021	11/02/2021
							F12	Serious Injury/ Illness/ Impairment	A0203	Defective Device	11/02/2021	11/02/2021
							F1905	Device Revision or Replacement	A0203	Defective Device	11/02/2021	11/02/2021
10740	2021	All other reportable incidents	35852	Single-chamber implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
10741	2021	All other reportable incidents	36241	Pacemaker/defibrillator lead	CT1684	Cardiac pacemakers and associated devices	E0604	Cardiac Perforation	A25	No Apparent Adverse Event		
							E0717	Dyspnea	A25	No Apparent Adverse Event		

							E0752	Unspecified Respiratory Problem	A25	No Apparent Adverse Event		
							E233001	Chest Pain	A25	No Apparent Adverse Event		
							F08	Hospitalization or Prolonged Hospitalization	A25	No Apparent Adverse Event		
							F14	Prolonged Episode of Care	A25	No Apparent Adverse Event		
10742	2021	All other reportable incidents	35852	Single-chamber implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A070909	Oversensing	17/02/2021	17/02/2021
									A071302	Inappropriate/Inadequate Shock/Stimulation	17/02/2021	17/02/2021
									A090801	Signal Artifact/Noise	17/02/2021	17/02/2021
									A1301	Failure to Read Input Signal	17/02/2021	17/02/2021
							F08	Hospitalization or Prolonged Hospitalization	A070909	Oversensing	17/02/2021	17/02/2021
									A071302	Inappropriate/Inadequate Shock/Stimulation	17/02/2021	17/02/2021
									A090801	Signal Artifact/Noise	17/02/2021	17/02/2021
									A1301	Failure to Read Input Signal	17/02/2021	17/02/2021
							F12	Serious Injury/ Illness/ Impairment	A070909	Oversensing	17/02/2021	17/02/2021
									A071302	Inappropriate/Inadequate Shock/Stimulation	17/02/2021	17/02/2021
									A090801	Signal Artifact/Noise	17/02/2021	17/02/2021
									A1301	Failure to Read Input Signal	17/02/2021	17/02/2021
							F1905	Device Revision or Replacement	A070909	Oversensing	17/02/2021	17/02/2021
									A071302	Inappropriate/Inadequate Shock/Stimulation	17/02/2021	17/02/2021
									A090801	Signal Artifact/Noise	17/02/2021	17/02/2021
									A1301	Failure to Read Input Signal	17/02/2021	17/02/2021
10743	2021	All other reportable incidents	35852	Single-chamber implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem	23/03/2021	23/03/2021
									A070504	Premature Discharge of Battery	23/03/2021	23/03/2021
							F19	Surgical Intervention	A0705	Battery Problem	23/03/2021	23/03/2021
									A070504	Premature Discharge of Battery	23/03/2021	23/03/2021
10744	2021	All other reportable incidents	47263	Cardiac resynchronization therapy implantable pacemaker	CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A0908	Incorrect, Inadequate or Imprecise Result or Readings	16/02/2021	16/02/2021
							F08	Hospitalization or Prolonged Hospitalization	A0908	Incorrect, Inadequate or Imprecise Result or Readings	16/02/2021	16/02/2021
							F1903	Device Explantation	A0908	Incorrect, Inadequate or Imprecise Result or Readings	16/02/2021	16/02/2021
							F1905	Device Revision or Replacement	A0908	Incorrect, Inadequate or Imprecise Result or Readings	16/02/2021	16/02/2021
					CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A0908	Incorrect, Inadequate or Imprecise Result or Readings	16/02/2021	16/02/2021
							F08	Hospitalization or Prolonged Hospitalization	A0908	Incorrect, Inadequate or Imprecise Result or Readings	16/02/2021	16/02/2021
							F1903	Device Explantation	A0908	Incorrect, Inadequate or Imprecise Result or Readings	16/02/2021	16/02/2021
							F1905	Device Revision or Replacement	A0908	Incorrect, Inadequate or Imprecise Result or Readings	16/02/2021	16/02/2021
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A0908	Incorrect, Inadequate or Imprecise Result or Readings	16/02/2021	16/02/2021
							F08	Hospitalization or Prolonged Hospitalization	A0908	Incorrect, Inadequate or Imprecise Result or Readings	16/02/2021	16/02/2021
							F1903	Device Explantation	A0908	Incorrect, Inadequate or Imprecise Result or Readings	16/02/2021	16/02/2021
							F1905	Device Revision or Replacement	A0908	Incorrect, Inadequate or Imprecise Result or Readings	16/02/2021	16/02/2021
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A0908	Incorrect, Inadequate or Imprecise Result or Readings	16/02/2021	16/02/2021
							F08	Hospitalization or Prolonged Hospitalization	A0908	Incorrect, Inadequate or Imprecise Result or Readings	16/02/2021	16/02/2021
							F1903	Device Explantation	A0908	Incorrect, Inadequate or Imprecise Result or Readings	16/02/2021	16/02/2021
							F1905	Device Revision or Replacement	A0908	Incorrect, Inadequate or Imprecise Result or Readings	16/02/2021	16/02/2021
10745	2021	All other reportable incidents	47265	Dual-chamber implantable pacemaker, rate-responsive	CT141	Electrodes/Leads and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	18/04/2021	18/04/2021
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	18/04/2021	18/04/2021
							F19	Surgical Intervention	A070504	Premature Discharge of Battery	18/04/2021	18/04/2021
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	18/04/2021	18/04/2021
					CT145	Cardiovascular devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	18/04/2021	18/04/2021
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	18/04/2021	18/04/2021
							F19	Surgical Intervention	A070504	Premature Discharge of Battery	18/04/2021	18/04/2021
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	18/04/2021	18/04/2021
					CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	18/04/2021	18/04/2021
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	18/04/2021	18/04/2021
							F19	Surgical Intervention	A070504	Premature Discharge of Battery	18/04/2021	18/04/2021
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	18/04/2021	18/04/2021
					CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	18/04/2021	18/04/2021
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	18/04/2021	18/04/2021
							F19	Surgical Intervention	A070504	Premature Discharge of Battery	18/04/2021	18/04/2021
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	18/04/2021	18/04/2021
					CT233	Surgical	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	18/04/2021	18/04/2021
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	18/04/2021	18/04/2021
							F19	Surgical Intervention	A070504	Premature Discharge of Battery	18/04/2021	18/04/2021
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	18/04/2021	18/04/2021
					CT2406	Implantable	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	18/04/2021	18/04/2021
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	18/04/2021	18/04/2021
							F19	Surgical Intervention	A070504	Premature Discharge of Battery	18/04/2021	18/04/2021
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	18/04/2021	18/04/2021
					CT2407	Active implantable	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	18/04/2021	18/04/2021
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	18/04/2021	18/04/2021
							F19	Surgical Intervention	A070504	Premature Discharge of Battery	18/04/2021	18/04/2021
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	18/04/2021	18/04/2021
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	18/04/2021	18/04/2021
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	18/04/2021	18/04/2021
							F19	Surgical Intervention	A070504	Premature Discharge of Battery	18/04/2021	18/04/2021
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	18/04/2021	18/04/2021
					CT2727	Active implant devices/ancillaries	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	18/04/2021	18/04/2021
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	18/04/2021	18/04/2021
							F19	Surgical Intervention	A070504	Premature Discharge of Battery	18/04/2021	18/04/2021
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	18/04/2021	18/04/2021

					CT2750	Implantable cardiac electrical stimulators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	18/04/2021	18/04/2021
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	18/04/2021	18/04/2021
							F19	Surgical Intervention	A070504	Premature Discharge of Battery	18/04/2021	18/04/2021
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	18/04/2021	18/04/2021
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	18/04/2021	18/04/2021
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	18/04/2021	18/04/2021
							F19	Surgical Intervention	A070504	Premature Discharge of Battery	18/04/2021	18/04/2021
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	18/04/2021	18/04/2021
					CT2985	Active	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	18/04/2021	18/04/2021
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	18/04/2021	18/04/2021
							F19	Surgical Intervention	A070504	Premature Discharge of Battery	18/04/2021	18/04/2021
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	18/04/2021	18/04/2021
					CT3001	Implantable metal	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	18/04/2021	18/04/2021
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	18/04/2021	18/04/2021
							F19	Surgical Intervention	A070504	Premature Discharge of Battery	18/04/2021	18/04/2021
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	18/04/2021	18/04/2021
					CT3002	Implantable synthetic polymer	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	18/04/2021	18/04/2021
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	18/04/2021	18/04/2021
							F19	Surgical Intervention	A070504	Premature Discharge of Battery	18/04/2021	18/04/2021
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	18/04/2021	18/04/2021
					CT3006	Implantable inorganic material	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	18/04/2021	18/04/2021
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	18/04/2021	18/04/2021
							F19	Surgical Intervention	A070504	Premature Discharge of Battery	18/04/2021	18/04/2021
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	18/04/2021	18/04/2021
					CT314	Electrically powered	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	18/04/2021	18/04/2021
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	18/04/2021	18/04/2021
							F19	Surgical Intervention	A070504	Premature Discharge of Battery	18/04/2021	18/04/2021
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	18/04/2021	18/04/2021
					CT321	Long-term surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	18/04/2021	18/04/2021
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	18/04/2021	18/04/2021
							F19	Surgical Intervention	A070504	Premature Discharge of Battery	18/04/2021	18/04/2021
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	18/04/2021	18/04/2021
					CT449	Generators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	18/04/2021	18/04/2021
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	18/04/2021	18/04/2021
							F19	Surgical Intervention	A070504	Premature Discharge of Battery	18/04/2021	18/04/2021
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	18/04/2021	18/04/2021
					CT981	Single use	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	18/04/2021	18/04/2021
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	18/04/2021	18/04/2021
							F19	Surgical Intervention	A070504	Premature Discharge of Battery	18/04/2021	18/04/2021
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	18/04/2021	18/04/2021
					CT983	Surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	18/04/2021	18/04/2021
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	18/04/2021	18/04/2021
							F19	Surgical Intervention	A070504	Premature Discharge of Battery	18/04/2021	18/04/2021
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	18/04/2021	18/04/2021
10746	2021	All other reportable incidents	35852	Single-chamber implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E0112	Dizziness	A070909	Oversensing		
									A070910	Undersensing		
									A071302	Inappropriate/Inadequate Shock/Stimulation		
							E0623	Unspecified Heart Problem	A070909	Oversensing		
									A070910	Undersensing		
									A071302	Inappropriate/Inadequate Shock/Stimulation		
							F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A070909	Oversensing		
									A070910	Undersensing		
									A071302	Inappropriate/Inadequate Shock/Stimulation		
							F26	No Health Consequences or Impact	A070909	Oversensing		
									A070910	Undersensing		
									A071302	Inappropriate/Inadequate Shock/Stimulation		
10747	2021	All other reportable incidents	35852	Single-chamber implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	22/02/2021	22/02/2021
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	22/02/2021	22/02/2021
							F12	Serious Injury/ Illness/ Impairment	A070504	Premature Discharge of Battery	22/02/2021	22/02/2021
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	22/02/2021	22/02/2021
10748	2021	All other reportable incidents	47265	Dual-chamber implantable pacemaker, rate-responsive	CT141	Electrodes/Leads and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	28/02/2021	28/02/2021
							F19	Surgical Intervention	A070504	Premature Discharge of Battery	28/02/2021	28/02/2021
					CT145	Cardiovascular devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	28/02/2021	28/02/2021
							F19	Surgical Intervention	A070504	Premature Discharge of Battery	28/02/2021	28/02/2021
					CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	28/02/2021	28/02/2021
							F19	Surgical Intervention	A070504	Premature Discharge of Battery	28/02/2021	28/02/2021
					CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	28/02/2021	28/02/2021
							F19	Surgical Intervention	A070504	Premature Discharge of Battery	28/02/2021	28/02/2021
					CT233	Surgical	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	28/02/2021	28/02/2021
							F19	Surgical Intervention	A070504	Premature Discharge of Battery	28/02/2021	28/02/2021
					CT2406	Implantable	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	28/02/2021	28/02/2021
							F19	Surgical Intervention	A070504	Premature Discharge of Battery	28/02/2021	28/02/2021
					CT2407	Active implantable	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	28/02/2021	28/02/2021
							F19	Surgical Intervention	A070504	Premature Discharge of Battery	28/02/2021	28/02/2021
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	28/02/2021	28/02/2021
							F19	Surgical Intervention	A070504	Premature Discharge of Battery	28/02/2021	28/02/2021
					CT2727	Active implant devices/ancillaries	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	28/02/2021	28/02/2021
							F19	Surgical Intervention	A070504	Premature Discharge of Battery	28/02/2021	28/02/2021
					CT2750	Implantable cardiac electrical stimulators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	28/02/2021	28/02/2021
							F19	Surgical Intervention	A070504	Premature Discharge of Battery	28/02/2021	28/02/2021
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	28/02/2021	28/02/2021
							F19	Surgical Intervention	A070504	Premature Discharge of Battery	28/02/2021	28/02/2021
					CT2985	Active	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	28/02/2021	28/02/2021
							F19	Surgical Intervention	A070504	Premature Discharge of Battery	28/02/2021	28/02/2021
					CT3001	Implantable metal	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	28/02/2021	28/02/2021
							F19	Surgical Intervention	A070504	Premature Discharge of Battery	28/02/2021	28/02/2021
					CT3002	Implantable synthetic polymer	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	28/02/2021	28/02/2021
							F19	Surgical Intervention	A070504	Premature Discharge of Battery	28/02/2021	28/02/2021

					CT3006	Implantable inorganic material	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	28/02/2021	28/02/2021
							F19	Surgical Intervention	A070504	Premature Discharge of Battery	28/02/2021	28/02/2021
					CT314	Electrically powered	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	28/02/2021	28/02/2021
							F19	Surgical Intervention	A070504	Premature Discharge of Battery	28/02/2021	28/02/2021
					CT321	Long-term surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	28/02/2021	28/02/2021
							F19	Surgical Intervention	A070504	Premature Discharge of Battery	28/02/2021	28/02/2021
					CT449	Generators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	28/02/2021	28/02/2021
							F19	Surgical Intervention	A070504	Premature Discharge of Battery	28/02/2021	28/02/2021
					CT981	Single use	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	28/02/2021	28/02/2021
							F19	Surgical Intervention	A070504	Premature Discharge of Battery	28/02/2021	28/02/2021
					CT983	Surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	28/02/2021	28/02/2021
							F19	Surgical Intervention	A070504	Premature Discharge of Battery	28/02/2021	28/02/2021
10749	2021	All other reportable incidents	47265	Dual-chamber implantable pacemaker, rate-responsive	CT141	Electrodes/Leads and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	13/04/2021	13/04/2021
							F19	Surgical Intervention	A070504	Premature Discharge of Battery	13/04/2021	13/04/2021
					CT145	Cardiovascular devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	13/04/2021	13/04/2021
							F19	Surgical Intervention	A070504	Premature Discharge of Battery	13/04/2021	13/04/2021
					CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	13/04/2021	13/04/2021
							F19	Surgical Intervention	A070504	Premature Discharge of Battery	13/04/2021	13/04/2021
					CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	13/04/2021	13/04/2021
							F19	Surgical Intervention	A070504	Premature Discharge of Battery	13/04/2021	13/04/2021
					CT233	Surgical	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	13/04/2021	13/04/2021
							F19	Surgical Intervention	A070504	Premature Discharge of Battery	13/04/2021	13/04/2021
					CT2406	Implantable	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	13/04/2021	13/04/2021
							F19	Surgical Intervention	A070504	Premature Discharge of Battery	13/04/2021	13/04/2021
					CT2407	Active implantable	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	13/04/2021	13/04/2021
							F19	Surgical Intervention	A070504	Premature Discharge of Battery	13/04/2021	13/04/2021
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	13/04/2021	13/04/2021
							F19	Surgical Intervention	A070504	Premature Discharge of Battery	13/04/2021	13/04/2021
					CT2727	Active implant devices/ancillaries	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	13/04/2021	13/04/2021
							F19	Surgical Intervention	A070504	Premature Discharge of Battery	13/04/2021	13/04/2021
					CT2750	Implantable cardiac electrical stimulators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	13/04/2021	13/04/2021
							F19	Surgical Intervention	A070504	Premature Discharge of Battery	13/04/2021	13/04/2021
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	13/04/2021	13/04/2021
							F19	Surgical Intervention	A070504	Premature Discharge of Battery	13/04/2021	13/04/2021
					CT2985	Active	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	13/04/2021	13/04/2021
							F19	Surgical Intervention	A070504	Premature Discharge of Battery	13/04/2021	13/04/2021
					CT3001	Implantable metal	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	13/04/2021	13/04/2021
							F19	Surgical Intervention	A070504	Premature Discharge of Battery	13/04/2021	13/04/2021
					CT3002	Implantable synthetic polymer	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	13/04/2021	13/04/2021
							F19	Surgical Intervention	A070504	Premature Discharge of Battery	13/04/2021	13/04/2021
					CT3006	Implantable inorganic material	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	13/04/2021	13/04/2021
							F19	Surgical Intervention	A070504	Premature Discharge of Battery	13/04/2021	13/04/2021
					CT314	Electrically powered	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	13/04/2021	13/04/2021
							F19	Surgical Intervention	A070504	Premature Discharge of Battery	13/04/2021	13/04/2021
					CT321	Long-term surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	13/04/2021	13/04/2021
							F19	Surgical Intervention	A070504	Premature Discharge of Battery	13/04/2021	13/04/2021
					CT449	Generators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	13/04/2021	13/04/2021
							F19	Surgical Intervention	A070504	Premature Discharge of Battery	13/04/2021	13/04/2021
					CT981	Single use	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	13/04/2021	13/04/2021
							F19	Surgical Intervention	A070504	Premature Discharge of Battery	13/04/2021	13/04/2021
					CT983	Surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	13/04/2021	13/04/2021
							F19	Surgical Intervention	A070504	Premature Discharge of Battery	13/04/2021	13/04/2021
10750	2021	All other reportable incidents	47265	Dual-chamber implantable pacemaker, rate-responsive	CT141	Electrodes/Leads and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	03/05/2021	03/05/2021
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	03/05/2021	03/05/2021
							F12	Serious Injury/ Illness/ Impairment	A070504	Premature Discharge of Battery	03/05/2021	03/05/2021
							F19	Surgical Intervention	A070504	Premature Discharge of Battery	03/05/2021	03/05/2021
					CT145	Cardiovascular devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	03/05/2021	03/05/2021
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	03/05/2021	03/05/2021
							F12	Serious Injury/ Illness/ Impairment	A070504	Premature Discharge of Battery	03/05/2021	03/05/2021
							F19	Surgical Intervention	A070504	Premature Discharge of Battery	03/05/2021	03/05/2021
					CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	03/05/2021	03/05/2021
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	03/05/2021	03/05/2021
							F12	Serious Injury/ Illness/ Impairment	A070504	Premature Discharge of Battery	03/05/2021	03/05/2021
							F19	Surgical Intervention	A070504	Premature Discharge of Battery	03/05/2021	03/05/2021
					CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	03/05/2021	03/05/2021
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	03/05/2021	03/05/2021
							F12	Serious Injury/ Illness/ Impairment	A070504	Premature Discharge of Battery	03/05/2021	03/05/2021
							F19	Surgical Intervention	A070504	Premature Discharge of Battery	03/05/2021	03/05/2021
					CT233	Surgical	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	03/05/2021	03/05/2021
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	03/05/2021	03/05/2021
							F12	Serious Injury/ Illness/ Impairment	A070504	Premature Discharge of Battery	03/05/2021	03/05/2021
							F19	Surgical Intervention	A070504	Premature Discharge of Battery	03/05/2021	03/05/2021
					CT2406	Implantable	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	03/05/2021	03/05/2021
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	03/05/2021	03/05/2021
							F12	Serious Injury/ Illness/ Impairment	A070504	Premature Discharge of Battery	03/05/2021	03/05/2021
							F19	Surgical Intervention	A070504	Premature Discharge of Battery	03/05/2021	03/05/2021
					CT2407	Active implantable	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	03/05/2021	03/05/2021
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	03/05/2021	03/05/2021
							F12	Serious Injury/ Illness/ Impairment	A070504	Premature Discharge of Battery	03/05/2021	03/05/2021
							F19	Surgical Intervention	A070504	Premature Discharge of Battery	03/05/2021	03/05/2021
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	03/05/2021	03/05/2021
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	03/05/2021	03/05/2021
							F12	Serious Injury/ Illness/ Impairment	A070504	Premature Discharge of Battery	03/05/2021	03/05/2021
							F19	Surgical Intervention	A070504	Premature Discharge of Battery	03/05/2021	03/05/2021
					CT2727	Active implant devices/ancillaries	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	03/05/2021	03/05/2021
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	03/05/2021	03/05/2021
							F12	Serious Injury/ Illness/ Impairment	A070504	Premature Discharge of Battery	03/05/2021	03/05/2021
							F19	Surgical Intervention	A070504	Premature Discharge of Battery	03/05/2021	03/05/2021

					CT2750	Implantable cardiac electrical stimulators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	03/05/2021	03/05/2021
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	03/05/2021	03/05/2021
							F12	Serious Injury/ Illness/ Impairment	A070504	Premature Discharge of Battery	03/05/2021	03/05/2021
							F19	Surgical Intervention	A070504	Premature Discharge of Battery	03/05/2021	03/05/2021
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	03/05/2021	03/05/2021
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	03/05/2021	03/05/2021
							F12	Serious Injury/ Illness/ Impairment	A070504	Premature Discharge of Battery	03/05/2021	03/05/2021
							F19	Surgical Intervention	A070504	Premature Discharge of Battery	03/05/2021	03/05/2021
					CT2985	Active	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	03/05/2021	03/05/2021
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	03/05/2021	03/05/2021
							F12	Serious Injury/ Illness/ Impairment	A070504	Premature Discharge of Battery	03/05/2021	03/05/2021
							F19	Surgical Intervention	A070504	Premature Discharge of Battery	03/05/2021	03/05/2021
					CT3001	Implantable metal	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	03/05/2021	03/05/2021
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	03/05/2021	03/05/2021
							F12	Serious Injury/ Illness/ Impairment	A070504	Premature Discharge of Battery	03/05/2021	03/05/2021
							F19	Surgical Intervention	A070504	Premature Discharge of Battery	03/05/2021	03/05/2021
					CT3002	Implantable synthetic polymer	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	03/05/2021	03/05/2021
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	03/05/2021	03/05/2021
							F12	Serious Injury/ Illness/ Impairment	A070504	Premature Discharge of Battery	03/05/2021	03/05/2021
							F19	Surgical Intervention	A070504	Premature Discharge of Battery	03/05/2021	03/05/2021
					CT3006	Implantable inorganic material	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	03/05/2021	03/05/2021
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	03/05/2021	03/05/2021
							F12	Serious Injury/ Illness/ Impairment	A070504	Premature Discharge of Battery	03/05/2021	03/05/2021
							F19	Surgical Intervention	A070504	Premature Discharge of Battery	03/05/2021	03/05/2021
					CT314	Electrically powered	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	03/05/2021	03/05/2021
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	03/05/2021	03/05/2021
							F12	Serious Injury/ Illness/ Impairment	A070504	Premature Discharge of Battery	03/05/2021	03/05/2021
							F19	Surgical Intervention	A070504	Premature Discharge of Battery	03/05/2021	03/05/2021
					CT321	Long-term surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	03/05/2021	03/05/2021
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	03/05/2021	03/05/2021
							F12	Serious Injury/ Illness/ Impairment	A070504	Premature Discharge of Battery	03/05/2021	03/05/2021
							F19	Surgical Intervention	A070504	Premature Discharge of Battery	03/05/2021	03/05/2021
					CT449	Generators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	03/05/2021	03/05/2021
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	03/05/2021	03/05/2021
							F12	Serious Injury/ Illness/ Impairment	A070504	Premature Discharge of Battery	03/05/2021	03/05/2021
							F19	Surgical Intervention	A070504	Premature Discharge of Battery	03/05/2021	03/05/2021
					CT981	Single use	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	03/05/2021	03/05/2021
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	03/05/2021	03/05/2021
							F12	Serious Injury/ Illness/ Impairment	A070504	Premature Discharge of Battery	03/05/2021	03/05/2021
							F19	Surgical Intervention	A070504	Premature Discharge of Battery	03/05/2021	03/05/2021
					CT983	Surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	03/05/2021	03/05/2021
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	03/05/2021	03/05/2021
							F12	Serious Injury/ Illness/ Impairment	A070504	Premature Discharge of Battery	03/05/2021	03/05/2021
							F19	Surgical Intervention	A070504	Premature Discharge of Battery	03/05/2021	03/05/2021
10751	2021	All other reportable incidents	36241	Pacemaker/defibrillator lead	CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A040101	Fracture		
									A070909	Oversensing		
									A071302	Inappropriate/Inadequate Shock/Stimulation		
							F26	No Health Consequences or Impact	A040101	Fracture		
									A070909	Oversensing		
									A071302	Inappropriate/Inadequate Shock/Stimulation		
10752	2021	All other reportable incidents	47265	Dual-chamber implantable pacemaker, rate-responsive	CT141	Electrodes/Leads and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery		
							F12	Serious Injury/ Illness/ Impairment	A070504	Premature Discharge of Battery		
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery		
					CT145	Cardiovascular devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery		
							F12	Serious Injury/ Illness/ Impairment	A070504	Premature Discharge of Battery		
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery		
					CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery		
							F12	Serious Injury/ Illness/ Impairment	A070504	Premature Discharge of Battery		
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery		
					CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery		
							F12	Serious Injury/ Illness/ Impairment	A070504	Premature Discharge of Battery		
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery		
					CT233	Surgical	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery		
							F12	Serious Injury/ Illness/ Impairment	A070504	Premature Discharge of Battery		
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery		
					CT2406	Implantable	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery		
							F12	Serious Injury/ Illness/ Impairment	A070504	Premature Discharge of Battery		
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery		
					CT2407	Active implantable	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery		
							F12	Serious Injury/ Illness/ Impairment	A070504	Premature Discharge of Battery		
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery		
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery		
							F12	Serious Injury/ Illness/ Impairment	A070504	Premature Discharge of Battery		
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery		
					CT2727	Active implant devices/ancillaries	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery		
							F12	Serious Injury/ Illness/ Impairment	A070504	Premature Discharge of Battery		
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery		
					CT2750	Implantable cardiac electrical stimulators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery		

							F12	Serious Injury/ Illness/ Impairment	A070504	Premature Discharge of Battery		
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery		
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery		
							F12	Serious Injury/ Illness/ Impairment	A070504	Premature Discharge of Battery		
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery		
					CT2985	Active	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery		
							F12	Serious Injury/ Illness/ Impairment	A070504	Premature Discharge of Battery		
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery		
					CT3001	Implantable metal	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery		
							F12	Serious Injury/ Illness/ Impairment	A070504	Premature Discharge of Battery		
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery		
					CT3002	Implantable synthetic polymer	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery		
							F12	Serious Injury/ Illness/ Impairment	A070504	Premature Discharge of Battery		
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery		
					CT3006	Implantable inorganic material	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery		
							F12	Serious Injury/ Illness/ Impairment	A070504	Premature Discharge of Battery		
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery		
					CT314	Electrically powered	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery		
							F12	Serious Injury/ Illness/ Impairment	A070504	Premature Discharge of Battery		
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery		
					CT321	Long-term surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery		
							F12	Serious Injury/ Illness/ Impairment	A070504	Premature Discharge of Battery		
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery		
					CT449	Generators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery		
							F12	Serious Injury/ Illness/ Impairment	A070504	Premature Discharge of Battery		
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery		
					CT981	Single use	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery		
							F12	Serious Injury/ Illness/ Impairment	A070504	Premature Discharge of Battery		
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery		
					CT983	Surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery		
							F12	Serious Injury/ Illness/ Impairment	A070504	Premature Discharge of Battery		
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery		
10753	2021	All other reportable incidents	35852	Single-chamber implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	21/03/2021	21/03/2021
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	21/03/2021	21/03/2021
							F12	Serious Injury/ Illness/ Impairment	A070504	Premature Discharge of Battery	21/03/2021	21/03/2021
							F19	Surgical Intervention	A070504	Premature Discharge of Battery	21/03/2021	21/03/2021
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	21/03/2021	21/03/2021
10754	2021	All other reportable incidents	47265	Dual-chamber implantable pacemaker, rate-responsive	CT141	Electrodes/Leads and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	17/03/2021	17/03/2021
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	17/03/2021	17/03/2021
							F12	Serious Injury/ Illness/ Impairment	A070504	Premature Discharge of Battery	17/03/2021	17/03/2021
							F19	Surgical Intervention	A070504	Premature Discharge of Battery	17/03/2021	17/03/2021
					CT145	Cardiovascular devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	17/03/2021	17/03/2021
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	17/03/2021	17/03/2021
							F12	Serious Injury/ Illness/ Impairment	A070504	Premature Discharge of Battery	17/03/2021	17/03/2021
							F19	Surgical Intervention	A070504	Premature Discharge of Battery	17/03/2021	17/03/2021
					CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	17/03/2021	17/03/2021
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	17/03/2021	17/03/2021
							F12	Serious Injury/ Illness/ Impairment	A070504	Premature Discharge of Battery	17/03/2021	17/03/2021
							F19	Surgical Intervention	A070504	Premature Discharge of Battery	17/03/2021	17/03/2021
					CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	17/03/2021	17/03/2021
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	17/03/2021	17/03/2021
							F12	Serious Injury/ Illness/ Impairment	A070504	Premature Discharge of Battery	17/03/2021	17/03/2021
							F19	Surgical Intervention	A070504	Premature Discharge of Battery	17/03/2021	17/03/2021
					CT233	Surgical	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	17/03/2021	17/03/2021
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	17/03/2021	17/03/2021
							F12	Serious Injury/ Illness/ Impairment	A070504	Premature Discharge of Battery	17/03/2021	17/03/2021
							F19	Surgical Intervention	A070504	Premature Discharge of Battery	17/03/2021	17/03/2021
					CT2406	Implantable	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	17/03/2021	17/03/2021
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	17/03/2021	17/03/2021
							F12	Serious Injury/ Illness/ Impairment	A070504	Premature Discharge of Battery	17/03/2021	17/03/2021
							F19	Surgical Intervention	A070504	Premature Discharge of Battery	17/03/2021	17/03/2021
					CT2407	Active implantable	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	17/03/2021	17/03/2021
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	17/03/2021	17/03/2021
							F12	Serious Injury/ Illness/ Impairment	A070504	Premature Discharge of Battery	17/03/2021	17/03/2021
							F19	Surgical Intervention	A070504	Premature Discharge of Battery	17/03/2021	17/03/2021
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	17/03/2021	17/03/2021
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	17/03/2021	17/03/2021
							F12	Serious Injury/ Illness/ Impairment	A070504	Premature Discharge of Battery	17/03/2021	17/03/2021
							F19	Surgical Intervention	A070504	Premature Discharge of Battery	17/03/2021	17/03/2021
					CT2727	Active implant devices/ancillaries	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	17/03/2021	17/03/2021
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	17/03/2021	17/03/2021
							F12	Serious Injury/ Illness/ Impairment	A070504	Premature Discharge of Battery	17/03/2021	17/03/2021
							F19	Surgical Intervention	A070504	Premature Discharge of Battery	17/03/2021	17/03/2021
					CT2750	Implantable cardiac electrical stimulators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	17/03/2021	17/03/2021
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	17/03/2021	17/03/2021
							F12	Serious Injury/ Illness/ Impairment	A070504	Premature Discharge of Battery	17/03/2021	17/03/2021
							F19	Surgical Intervention	A070504	Premature Discharge of Battery	17/03/2021	17/03/2021
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	17/03/2021	17/03/2021

							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	17/03/2021	17/03/2021
							F12	Serious Injury/ Illness/ Impairment	A070504	Premature Discharge of Battery	17/03/2021	17/03/2021
							F19	Surgical Intervention	A070504	Premature Discharge of Battery	17/03/2021	17/03/2021
					CT2985	Active	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	17/03/2021	17/03/2021
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	17/03/2021	17/03/2021
							F12	Serious Injury/ Illness/ Impairment	A070504	Premature Discharge of Battery	17/03/2021	17/03/2021
							F19	Surgical Intervention	A070504	Premature Discharge of Battery	17/03/2021	17/03/2021
					CT3001	Implantable metal	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	17/03/2021	17/03/2021
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	17/03/2021	17/03/2021
							F12	Serious Injury/ Illness/ Impairment	A070504	Premature Discharge of Battery	17/03/2021	17/03/2021
							F19	Surgical Intervention	A070504	Premature Discharge of Battery	17/03/2021	17/03/2021
					CT3002	Implantable synthetic polymer	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	17/03/2021	17/03/2021
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	17/03/2021	17/03/2021
							F12	Serious Injury/ Illness/ Impairment	A070504	Premature Discharge of Battery	17/03/2021	17/03/2021
							F19	Surgical Intervention	A070504	Premature Discharge of Battery	17/03/2021	17/03/2021
					CT3006	Implantable inorganic material	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	17/03/2021	17/03/2021
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	17/03/2021	17/03/2021
							F12	Serious Injury/ Illness/ Impairment	A070504	Premature Discharge of Battery	17/03/2021	17/03/2021
							F19	Surgical Intervention	A070504	Premature Discharge of Battery	17/03/2021	17/03/2021
					CT314	Electrically powered	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	17/03/2021	17/03/2021
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	17/03/2021	17/03/2021
							F12	Serious Injury/ Illness/ Impairment	A070504	Premature Discharge of Battery	17/03/2021	17/03/2021
							F19	Surgical Intervention	A070504	Premature Discharge of Battery	17/03/2021	17/03/2021
					CT321	Long-term surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	17/03/2021	17/03/2021
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	17/03/2021	17/03/2021
							F12	Serious Injury/ Illness/ Impairment	A070504	Premature Discharge of Battery	17/03/2021	17/03/2021
							F19	Surgical Intervention	A070504	Premature Discharge of Battery	17/03/2021	17/03/2021
					CT449	Generators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	17/03/2021	17/03/2021
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	17/03/2021	17/03/2021
							F12	Serious Injury/ Illness/ Impairment	A070504	Premature Discharge of Battery	17/03/2021	17/03/2021
							F19	Surgical Intervention	A070504	Premature Discharge of Battery	17/03/2021	17/03/2021
					CT981	Single use	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	17/03/2021	17/03/2021
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	17/03/2021	17/03/2021
							F12	Serious Injury/ Illness/ Impairment	A070504	Premature Discharge of Battery	17/03/2021	17/03/2021
							F19	Surgical Intervention	A070504	Premature Discharge of Battery	17/03/2021	17/03/2021
					CT983	Surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	17/03/2021	17/03/2021
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	17/03/2021	17/03/2021
							F12	Serious Injury/ Illness/ Impairment	A070504	Premature Discharge of Battery	17/03/2021	17/03/2021
							F19	Surgical Intervention	A070504	Premature Discharge of Battery	17/03/2021	17/03/2021
10755	2021	All other reportable incidents	17882	Physiologic-monitoring defibrillation system	CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070908	Failure to Sense		
							F26	No Health Consequences or Impact	A070908	Failure to Sense		
10756	2021	All other reportable incidents	35852	Single-chamber implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E2007	Fail	A070909	Oversensing		
							E2403	No Clinical Signs, Symptoms or Conditions	A070909	Oversensing		
							F26	No Health Consequences or Impact	A070909	Oversensing		
10757	2021	All other reportable incidents	47265	Dual-chamber implantable pacemaker, rate-responsive	CT141	Electrodes/Leads and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	01/05/2021	01/05/2021
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	01/05/2021	01/05/2021
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	01/05/2021	01/05/2021
					CT145	Cardiovascular devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	01/05/2021	01/05/2021
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	01/05/2021	01/05/2021
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	01/05/2021	01/05/2021
					CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	01/05/2021	01/05/2021
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	01/05/2021	01/05/2021
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	01/05/2021	01/05/2021
					CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	01/05/2021	01/05/2021
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	01/05/2021	01/05/2021
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	01/05/2021	01/05/2021
					CT233	Surgical	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	01/05/2021	01/05/2021
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	01/05/2021	01/05/2021
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	01/05/2021	01/05/2021
					CT2406	Implantable	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	01/05/2021	01/05/2021
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	01/05/2021	01/05/2021
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	01/05/2021	01/05/2021
					CT2407	Active implantable	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	01/05/2021	01/05/2021
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	01/05/2021	01/05/2021
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	01/05/2021	01/05/2021
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	01/05/2021	01/05/2021
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	01/05/2021	01/05/2021
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	01/05/2021	01/05/2021
					CT2727	Active implant devices/ancillaries	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	01/05/2021	01/05/2021
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	01/05/2021	01/05/2021
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	01/05/2021	01/05/2021
					CT2750	Implantable cardiac electrical stimulators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	01/05/2021	01/05/2021
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	01/05/2021	01/05/2021
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	01/05/2021	01/05/2021
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	01/05/2021	01/05/2021
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	01/05/2021	01/05/2021
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	01/05/2021	01/05/2021
					CT2985	Active	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	01/05/2021	01/05/2021
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	01/05/2021	01/05/2021
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	01/05/2021	01/05/2021
					CT3001	Implantable metal	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	01/05/2021	01/05/2021
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	01/05/2021	01/05/2021
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	01/05/2021	01/05/2021
					CT3002	Implantable synthetic polymer	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	01/05/2021	01/05/2021
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	01/05/2021	01/05/2021
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	01/05/2021	01/05/2021
					CT3006	Implantable inorganic material	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	01/05/2021	01/05/2021
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	01/05/2021	01/05/2021

							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	01/05/2021	01/05/2021
					CT314	Electrically powered	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	01/05/2021	01/05/2021
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	01/05/2021	01/05/2021
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	01/05/2021	01/05/2021
					CT321	Long-term surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	01/05/2021	01/05/2021
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	01/05/2021	01/05/2021
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	01/05/2021	01/05/2021
					CT449	Generators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	01/05/2021	01/05/2021
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	01/05/2021	01/05/2021
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	01/05/2021	01/05/2021
					CT981	Single use	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	01/05/2021	01/05/2021
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	01/05/2021	01/05/2021
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	01/05/2021	01/05/2021
					CT983	Surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	01/05/2021	01/05/2021
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	01/05/2021	01/05/2021
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	01/05/2021	01/05/2021
10758	2021	All other reportable incidents	35852	Single-chamber implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E2401	Insufficient Information	A1304	Telemetry Discrepancy		
							E2403	No Clinical Signs, Symptoms or Conditions	A1304	Telemetry Discrepancy		
							F26	No Health Consequences or Impact	A1304	Telemetry Discrepancy		
10759	2021	All other reportable incidents	35223	Endocardial/interventricular septal pacing lead	CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070909	Oversensing		
									A090801	Signal Artifact/Noise		
							F26	No Health Consequences or Impact	A070909	Oversensing		
									A090801	Signal Artifact/Noise		
10760	2021	All other reportable incidents	35852	Single-chamber implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E2104	Electric Shock	A071302	Inappropriate/Inadequate Shock/Stimulation		
							F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A071302	Inappropriate/Inadequate Shock/Stimulation		
							F26	No Health Consequences or Impact	A071302	Inappropriate/Inadequate Shock/Stimulation		
10761	2021	All other reportable incidents	47270	Cardiac resynchronization therapy implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E060109	Tachycardia	A070910	Undersensing		
									A0713	Defibrillation/Stimulation Problem		
							F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A070910	Undersensing		
									A0713	Defibrillation/Stimulation Problem		
							F26	No Health Consequences or Impact	A070910	Undersensing		
									A0713	Defibrillation/Stimulation Problem		
10762	2021	Death	35852	Single-chamber implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E060109	Tachycardia	A0712	Pacing Problem	01/03/2021	01/03/2021
							E2401	Insufficient Information	A0712	Pacing Problem	01/03/2021	01/03/2021
							F02	Death	A0712	Pacing Problem	01/03/2021	01/03/2021
10763	2021	All other reportable incidents	47265	Dual-chamber implantable pacemaker, rate-responsive	CT141	Electrodes/Leads and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	26/05/2021	26/05/2021
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	26/05/2021	26/05/2021
							F1903	Device Explantation	A070504	Premature Discharge of Battery	26/05/2021	26/05/2021
					CT145	Cardiovascular devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	26/05/2021	26/05/2021
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	26/05/2021	26/05/2021
							F1903	Device Explantation	A070504	Premature Discharge of Battery	26/05/2021	26/05/2021
					CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	26/05/2021	26/05/2021
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	26/05/2021	26/05/2021
							F1903	Device Explantation	A070504	Premature Discharge of Battery	26/05/2021	26/05/2021
					CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	26/05/2021	26/05/2021
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	26/05/2021	26/05/2021
							F1903	Device Explantation	A070504	Premature Discharge of Battery	26/05/2021	26/05/2021
					CT233	Surgical	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	26/05/2021	26/05/2021
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	26/05/2021	26/05/2021
							F1903	Device Explantation	A070504	Premature Discharge of Battery	26/05/2021	26/05/2021
					CT2406	Implantable	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	26/05/2021	26/05/2021
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	26/05/2021	26/05/2021
							F1903	Device Explantation	A070504	Premature Discharge of Battery	26/05/2021	26/05/2021
					CT2407	Active implantable	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	26/05/2021	26/05/2021
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	26/05/2021	26/05/2021
							F1903	Device Explantation	A070504	Premature Discharge of Battery	26/05/2021	26/05/2021
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	26/05/2021	26/05/2021
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	26/05/2021	26/05/2021
							F1903	Device Explantation	A070504	Premature Discharge of Battery	26/05/2021	26/05/2021
					CT2727	Active implant devices/ancillaries	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	26/05/2021	26/05/2021
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	26/05/2021	26/05/2021
							F1903	Device Explantation	A070504	Premature Discharge of Battery	26/05/2021	26/05/2021
					CT2750	Implantable cardiac electrical stimulators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	26/05/2021	26/05/2021
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	26/05/2021	26/05/2021
							F1903	Device Explantation	A070504	Premature Discharge of Battery	26/05/2021	26/05/2021
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	26/05/2021	26/05/2021
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	26/05/2021	26/05/2021
							F1903	Device Explantation	A070504	Premature Discharge of Battery	26/05/2021	26/05/2021
					CT2985	Active	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	26/05/2021	26/05/2021
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	26/05/2021	26/05/2021
							F1903	Device Explantation	A070504	Premature Discharge of Battery	26/05/2021	26/05/2021
					CT3001	Implantable metal	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	26/05/2021	26/05/2021
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	26/05/2021	26/05/2021
							F1903	Device Explantation	A070504	Premature Discharge of Battery	26/05/2021	26/05/2021
					CT3002	Implantable synthetic polymer	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	26/05/2021	26/05/2021
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	26/05/2021	26/05/2021
							F1903	Device Explantation	A070504	Premature Discharge of Battery	26/05/2021	26/05/2021
					CT3006	Implantable inorganic material	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	26/05/2021	26/05/2021
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	26/05/2021	26/05/2021
							F1903	Device Explantation	A070504	Premature Discharge of Battery	26/05/2021	26/05/2021
					CT314	Electrically powered	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	26/05/2021	26/05/2021
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	26/05/2021	26/05/2021
							F1903	Device Explantation	A070504	Premature Discharge of Battery	26/05/2021	26/05/2021
					CT321	Long-term surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	26/05/2021	26/05/2021
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	26/05/2021	26/05/2021
							F1903	Device Explantation	A070504	Premature Discharge of Battery	26/05/2021	26/05/2021
					CT449	Generators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	26/05/2021	26/05/2021
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	26/05/2021	26/05/2021

						F1903	Device Explantation	A070504	Premature Discharge of Battery	26/05/2021	26/05/2021
					CT981	Single use	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	26/05/2021	26/05/2021
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	26/05/2021	26/05/2021
							F1903	Device Explantation	A070504	Premature Discharge of Battery	26/05/2021	26/05/2021
					CT983	Surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	26/05/2021	26/05/2021
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	26/05/2021	26/05/2021
							F1903	Device Explantation	A070504	Premature Discharge of Battery	26/05/2021	26/05/2021
10764	2021	All other reportable incidents	47265	Dual-chamber implantable pacemaker, rate-responsive	CT141	Electrodes/Leads and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	03/03/2021	03/03/2021
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	03/03/2021	03/03/2021
							F12	Serious Injury/ Illness/ Impairment	A070504	Premature Discharge of Battery	03/03/2021	03/03/2021
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	03/03/2021	03/03/2021
					CT145	Cardiovascular devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	03/03/2021	03/03/2021
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	03/03/2021	03/03/2021
							F12	Serious Injury/ Illness/ Impairment	A070504	Premature Discharge of Battery	03/03/2021	03/03/2021
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	03/03/2021	03/03/2021
					CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	03/03/2021	03/03/2021
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	03/03/2021	03/03/2021
							F12	Serious Injury/ Illness/ Impairment	A070504	Premature Discharge of Battery	03/03/2021	03/03/2021
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	03/03/2021	03/03/2021
					CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	03/03/2021	03/03/2021
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	03/03/2021	03/03/2021
							F12	Serious Injury/ Illness/ Impairment	A070504	Premature Discharge of Battery	03/03/2021	03/03/2021
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	03/03/2021	03/03/2021
					CT233	Surgical	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	03/03/2021	03/03/2021
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	03/03/2021	03/03/2021
							F12	Serious Injury/ Illness/ Impairment	A070504	Premature Discharge of Battery	03/03/2021	03/03/2021
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	03/03/2021	03/03/2021
					CT2406	Implantable	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	03/03/2021	03/03/2021
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	03/03/2021	03/03/2021
							F12	Serious Injury/ Illness/ Impairment	A070504	Premature Discharge of Battery	03/03/2021	03/03/2021
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	03/03/2021	03/03/2021
					CT2407	Active implantable	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	03/03/2021	03/03/2021
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	03/03/2021	03/03/2021
							F12	Serious Injury/ Illness/ Impairment	A070504	Premature Discharge of Battery	03/03/2021	03/03/2021
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	03/03/2021	03/03/2021
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	03/03/2021	03/03/2021
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	03/03/2021	03/03/2021
							F12	Serious Injury/ Illness/ Impairment	A070504	Premature Discharge of Battery	03/03/2021	03/03/2021
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	03/03/2021	03/03/2021
					CT2727	Active implant devices/ancillaries	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	03/03/2021	03/03/2021
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	03/03/2021	03/03/2021
							F12	Serious Injury/ Illness/ Impairment	A070504	Premature Discharge of Battery	03/03/2021	03/03/2021
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	03/03/2021	03/03/2021
					CT2750	Implantable cardiac electrical stimulators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	03/03/2021	03/03/2021
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	03/03/2021	03/03/2021
							F12	Serious Injury/ Illness/ Impairment	A070504	Premature Discharge of Battery	03/03/2021	03/03/2021
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	03/03/2021	03/03/2021
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	03/03/2021	03/03/2021
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	03/03/2021	03/03/2021
							F12	Serious Injury/ Illness/ Impairment	A070504	Premature Discharge of Battery	03/03/2021	03/03/2021
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	03/03/2021	03/03/2021
					CT2985	Active	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	03/03/2021	03/03/2021
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	03/03/2021	03/03/2021
							F12	Serious Injury/ Illness/ Impairment	A070504	Premature Discharge of Battery	03/03/2021	03/03/2021
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	03/03/2021	03/03/2021
					CT3001	Implantable metal	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	03/03/2021	03/03/2021
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	03/03/2021	03/03/2021
							F12	Serious Injury/ Illness/ Impairment	A070504	Premature Discharge of Battery	03/03/2021	03/03/2021
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	03/03/2021	03/03/2021
					CT3002	Implantable synthetic polymer	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	03/03/2021	03/03/2021
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	03/03/2021	03/03/2021
							F12	Serious Injury/ Illness/ Impairment	A070504	Premature Discharge of Battery	03/03/2021	03/03/2021
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	03/03/2021	03/03/2021
					CT3006	Implantable inorganic material	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	03/03/2021	03/03/2021
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	03/03/2021	03/03/2021
							F12	Serious Injury/ Illness/ Impairment	A070504	Premature Discharge of Battery	03/03/2021	03/03/2021
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	03/03/2021	03/03/2021
					CT314	Electrically powered	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	03/03/2021	03/03/2021
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	03/03/2021	03/03/2021
							F12	Serious Injury/ Illness/ Impairment	A070504	Premature Discharge of Battery	03/03/2021	03/03/2021
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	03/03/2021	03/03/2021
					CT321	Long-term surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	03/03/2021	03/03/2021
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	03/03/2021	03/03/2021
							F12	Serious Injury/ Illness/ Impairment	A070504	Premature Discharge of Battery	03/03/2021	03/03/2021
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	03/03/2021	03/03/2021
					CT449	Generators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	03/03/2021	03/03/2021
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	03/03/2021	03/03/2021
							F12	Serious Injury/ Illness/ Impairment	A070504	Premature Discharge of Battery	03/03/2021	03/03/2021
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	03/03/2021	03/03/2021
					CT981	Single use	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	03/03/2021	03/03/2021
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	03/03/2021	03/03/2021
							F12	Serious Injury/ Illness/ Impairment	A070504	Premature Discharge of Battery	03/03/2021	03/03/2021
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	03/03/2021	03/03/2021
					CT983	Surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	03/03/2021	03/03/2021
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	03/03/2021	03/03/2021
							F12	Serious Injury/ Illness/ Impairment	A070504	Premature Discharge of Battery	03/03/2021	03/03/2021
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	03/03/2021	03/03/2021
10765	2021	All other reportable incidents	47265	Dual-chamber implantable pacemaker, rate-	CT141	Electrodes/Leads and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A072201	High Impedance		

				responsive			F26	No Health Consequences or Impact	A072201	High Impedance		
					CT145	Cardiovascular devices	E2403	No Clinical Signs, Symptoms or Conditions	A072201	High Impedance		
							F26	No Health Consequences or Impact	A072201	High Impedance		
					CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A072201	High Impedance		
							F26	No Health Consequences or Impact	A072201	High Impedance		
					CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A072201	High Impedance		
							F26	No Health Consequences or Impact	A072201	High Impedance		
					CT233	Surgical	E2403	No Clinical Signs, Symptoms or Conditions	A072201	High Impedance		
							F26	No Health Consequences or Impact	A072201	High Impedance		
					CT2406	Implantable	E2403	No Clinical Signs, Symptoms or Conditions	A072201	High Impedance		
							F26	No Health Consequences or Impact	A072201	High Impedance		
					CT2407	Active implantable	E2403	No Clinical Signs, Symptoms or Conditions	A072201	High Impedance		
							F26	No Health Consequences or Impact	A072201	High Impedance		
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A072201	High Impedance		
							F26	No Health Consequences or Impact	A072201	High Impedance		
					CT2727	Active implant devices/ancillaries	E2403	No Clinical Signs, Symptoms or Conditions	A072201	High Impedance		
							F26	No Health Consequences or Impact	A072201	High Impedance		
					CT2750	Implantable cardiac electrical stimulators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A072201	High Impedance		
							F26	No Health Consequences or Impact	A072201	High Impedance		
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A072201	High Impedance		
							F26	No Health Consequences or Impact	A072201	High Impedance		
					CT2985	Active	E2403	No Clinical Signs, Symptoms or Conditions	A072201	High Impedance		
							F26	No Health Consequences or Impact	A072201	High Impedance		
					CT3001	Implantable metal	E2403	No Clinical Signs, Symptoms or Conditions	A072201	High Impedance		
							F26	No Health Consequences or Impact	A072201	High Impedance		
					CT3002	Implantable synthetic polymer	E2403	No Clinical Signs, Symptoms or Conditions	A072201	High Impedance		
							F26	No Health Consequences or Impact	A072201	High Impedance		
					CT3006	Implantable inorganic material	E2403	No Clinical Signs, Symptoms or Conditions	A072201	High Impedance		
							F26	No Health Consequences or Impact	A072201	High Impedance		
					CT314	Electrically powered	E2403	No Clinical Signs, Symptoms or Conditions	A072201	High Impedance		
							F26	No Health Consequences or Impact	A072201	High Impedance		
					CT321	Long-term surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A072201	High Impedance		
							F26	No Health Consequences or Impact	A072201	High Impedance		
					CT449	Generators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A072201	High Impedance		
							F26	No Health Consequences or Impact	A072201	High Impedance		
					CT981	Single use	E2403	No Clinical Signs, Symptoms or Conditions	A072201	High Impedance		
							F26	No Health Consequences or Impact	A072201	High Impedance		
					CT983	Surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A072201	High Impedance		
							F26	No Health Consequences or Impact	A072201	High Impedance		
10766	2021	All other reportable incidents	35852	Single-chamber implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem	19/02/2021	19/02/2021
									A070504	Premature Discharge of Battery	19/02/2021	19/02/2021
							F08	Hospitalization or Prolonged Hospitalization	A0705	Battery Problem	19/02/2021	19/02/2021
									A070504	Premature Discharge of Battery	19/02/2021	19/02/2021
							F1903	Device Explantation	A0705	Battery Problem	19/02/2021	19/02/2021
									A070504	Premature Discharge of Battery	19/02/2021	19/02/2021
							F1905	Device Revision or Replacement	A0705	Battery Problem	19/02/2021	19/02/2021
									A070504	Premature Discharge of Battery	19/02/2021	19/02/2021
10767	2021	All other reportable incidents	35852	Single-chamber implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	09/03/2021	09/03/2021
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	09/03/2021	09/03/2021
							F19	Surgical Intervention	A070504	Premature Discharge of Battery	09/03/2021	09/03/2021
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	09/03/2021	09/03/2021
10768	2021	All other reportable incidents	35852	Single-chamber implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E2104	Electric Shock	A071302	Inappropriate/Inadequate Shock/Stimulation		
							E2403	No Clinical Signs, Symptoms or Conditions	A071302	Inappropriate/Inadequate Shock/Stimulation		
							F26	No Health Consequences or Impact	A071302	Inappropriate/Inadequate Shock/Stimulation		
10769	2021	All other reportable incidents	47265	Dual-chamber implantable pacemaker, rate-responsive	CT141	Electrodes/Leads and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
					CT145	Cardiovascular devices	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
					CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
					CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
					CT233	Surgical	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
					CT2406	Implantable	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
					CT2407	Active implantable	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
					CT2727	Active implant devices/ancillaries	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		

							F26	No Health Consequences or Impact	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
					CT2750	Implantable cardiac electrical stimulators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
					CT2985	Active	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
					CT3001	Implantable metal	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
					CT3002	Implantable synthetic polymer	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
					CT3006	Implantable inorganic material	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
					CT314	Electrically powered	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
					CT321	Long-term surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
					CT449	Generators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
					CT981	Single use	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
					CT983	Surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
10770	2021	All other reportable incidents	47270	Cardiac resynchronization therapy implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A04	Material Integrity Problem	08/03/2021	08/03/2021
									A12	Connection Problem	08/03/2021	08/03/2021
							F19	Surgical Intervention	A04	Material Integrity Problem	08/03/2021	08/03/2021
									A12	Connection Problem	08/03/2021	08/03/2021
							F1905	Device Revision or Replacement	A04	Material Integrity Problem	08/03/2021	08/03/2021
									A12	Connection Problem	08/03/2021	08/03/2021
10771	2021	All other reportable incidents	47265	Dual-chamber implantable pacemaker, rate-responsive	CT141	Electrodes/Leads and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	19/04/2021	19/04/2021
							F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A070504	Premature Discharge of Battery	19/04/2021	19/04/2021
							F19	Surgical Intervention	A070504	Premature Discharge of Battery	19/04/2021	19/04/2021
					CT145	Cardiovascular devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	19/04/2021	19/04/2021
							F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A070504	Premature Discharge of Battery	19/04/2021	19/04/2021
							F19	Surgical Intervention	A070504	Premature Discharge of Battery	19/04/2021	19/04/2021
					CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	19/04/2021	19/04/2021
							F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A070504	Premature Discharge of Battery	19/04/2021	19/04/2021
							F19	Surgical Intervention	A070504	Premature Discharge of Battery	19/04/2021	19/04/2021
					CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	19/04/2021	19/04/2021
							F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A070504	Premature Discharge of Battery	19/04/2021	19/04/2021
							F19	Surgical Intervention	A070504	Premature Discharge of Battery	19/04/2021	19/04/2021
					CT233	Surgical	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	19/04/2021	19/04/2021
							F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A070504	Premature Discharge of Battery	19/04/2021	19/04/2021
							F19	Surgical Intervention	A070504	Premature Discharge of Battery	19/04/2021	19/04/2021
					CT2406	Implantable	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	19/04/2021	19/04/2021
							F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A070504	Premature Discharge of Battery	19/04/2021	19/04/2021
							F19	Surgical Intervention	A070504	Premature Discharge of Battery	19/04/2021	19/04/2021
					CT2407	Active implantable	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	19/04/2021	19/04/2021
							F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A070504	Premature Discharge of Battery	19/04/2021	19/04/2021
							F19	Surgical Intervention	A070504	Premature Discharge of Battery	19/04/2021	19/04/2021
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	19/04/2021	19/04/2021
							F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A070504	Premature Discharge of Battery	19/04/2021	19/04/2021
							F19	Surgical Intervention	A070504	Premature Discharge of Battery	19/04/2021	19/04/2021
					CT2727	Active implant devices/ancillaries	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	19/04/2021	19/04/2021
							F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A070504	Premature Discharge of Battery	19/04/2021	19/04/2021
							F19	Surgical Intervention	A070504	Premature Discharge of Battery	19/04/2021	19/04/2021
					CT2750	Implantable cardiac electrical stimulators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	19/04/2021	19/04/2021
							F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A070504	Premature Discharge of Battery	19/04/2021	19/04/2021
							F19	Surgical Intervention	A070504	Premature Discharge of Battery	19/04/2021	19/04/2021
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	19/04/2021	19/04/2021
							F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A070504	Premature Discharge of Battery	19/04/2021	19/04/2021
							F19	Surgical Intervention	A070504	Premature Discharge of Battery	19/04/2021	19/04/2021
					CT2985	Active	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	19/04/2021	19/04/2021
							F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A070504	Premature Discharge of Battery	19/04/2021	19/04/2021
							F19	Surgical Intervention	A070504	Premature Discharge of Battery	19/04/2021	19/04/2021

					CT3001	Implantable metal	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	19/04/2021	19/04/2021
							F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A070504	Premature Discharge of Battery	19/04/2021	19/04/2021
							F19	Surgical Intervention	A070504	Premature Discharge of Battery	19/04/2021	19/04/2021
					CT3002	Implantable synthetic polymer	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	19/04/2021	19/04/2021
							F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A070504	Premature Discharge of Battery	19/04/2021	19/04/2021
							F19	Surgical Intervention	A070504	Premature Discharge of Battery	19/04/2021	19/04/2021
					CT3006	Implantable inorganic material	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	19/04/2021	19/04/2021
							F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A070504	Premature Discharge of Battery	19/04/2021	19/04/2021
							F19	Surgical Intervention	A070504	Premature Discharge of Battery	19/04/2021	19/04/2021
					CT314	Electrically powered	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	19/04/2021	19/04/2021
							F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A070504	Premature Discharge of Battery	19/04/2021	19/04/2021
							F19	Surgical Intervention	A070504	Premature Discharge of Battery	19/04/2021	19/04/2021
					CT321	Long-term surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	19/04/2021	19/04/2021
							F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A070504	Premature Discharge of Battery	19/04/2021	19/04/2021
							F19	Surgical Intervention	A070504	Premature Discharge of Battery	19/04/2021	19/04/2021
					CT449	Generators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	19/04/2021	19/04/2021
							F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A070504	Premature Discharge of Battery	19/04/2021	19/04/2021
							F19	Surgical Intervention	A070504	Premature Discharge of Battery	19/04/2021	19/04/2021
					CT981	Single use	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	19/04/2021	19/04/2021
							F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A070504	Premature Discharge of Battery	19/04/2021	19/04/2021
							F19	Surgical Intervention	A070504	Premature Discharge of Battery	19/04/2021	19/04/2021
					CT983	Surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	19/04/2021	19/04/2021
							F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A070504	Premature Discharge of Battery	19/04/2021	19/04/2021
							F19	Surgical Intervention	A070504	Premature Discharge of Battery	19/04/2021	19/04/2021
10772	2021	All other reportable incidents	47267	Single-chamber implantable pacemaker, rate-responsive	CT141	Electrodes/Leads and associated devices	E011903	Syncope/Fainting	A071302	Inappropriate/Inadequate Shock/Stimulation		
							F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A071302	Inappropriate/Inadequate Shock/Stimulation		
							F26	No Health Consequences or Impact	A071302	Inappropriate/Inadequate Shock/Stimulation		
					CT145	Cardiovascular devices	E011903	Syncope/Fainting	A071302	Inappropriate/Inadequate Shock/Stimulation		
							F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A071302	Inappropriate/Inadequate Shock/Stimulation		
							F26	No Health Consequences or Impact	A071302	Inappropriate/Inadequate Shock/Stimulation		
					CT1684	Cardiac pacemakers and associated devices	E011903	Syncope/Fainting	A071302	Inappropriate/Inadequate Shock/Stimulation		
							F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A071302	Inappropriate/Inadequate Shock/Stimulation		
							F26	No Health Consequences or Impact	A071302	Inappropriate/Inadequate Shock/Stimulation		
					CT1685	Cardiac pacemakers	E011903	Syncope/Fainting	A071302	Inappropriate/Inadequate Shock/Stimulation		
							F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A071302	Inappropriate/Inadequate Shock/Stimulation		
							F26	No Health Consequences or Impact	A071302	Inappropriate/Inadequate Shock/Stimulation		
					CT233	Surgical	E011903	Syncope/Fainting	A071302	Inappropriate/Inadequate Shock/Stimulation		
							F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A071302	Inappropriate/Inadequate Shock/Stimulation		
							F26	No Health Consequences or Impact	A071302	Inappropriate/Inadequate Shock/Stimulation		
					CT2406	Implantable	E011903	Syncope/Fainting	A071302	Inappropriate/Inadequate Shock/Stimulation		
							F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A071302	Inappropriate/Inadequate Shock/Stimulation		
							F26	No Health Consequences or Impact	A071302	Inappropriate/Inadequate Shock/Stimulation		
					CT2407	Active implantable	E011903	Syncope/Fainting	A071302	Inappropriate/Inadequate Shock/Stimulation		
							F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A071302	Inappropriate/Inadequate Shock/Stimulation		
							F26	No Health Consequences or Impact	A071302	Inappropriate/Inadequate Shock/Stimulation		
					CT257	Implantable cardiac pacemakers	E011903	Syncope/Fainting	A071302	Inappropriate/Inadequate Shock/Stimulation		
							F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A071302	Inappropriate/Inadequate Shock/Stimulation		
							F26	No Health Consequences or Impact	A071302	Inappropriate/Inadequate Shock/Stimulation		
					CT2727	Active implant devices/ancillaries	E011903	Syncope/Fainting	A071302	Inappropriate/Inadequate Shock/Stimulation		
							F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A071302	Inappropriate/Inadequate Shock/Stimulation		
							F26	No Health Consequences or Impact	A071302	Inappropriate/Inadequate Shock/Stimulation		
					CT2750	Implantable cardiac electrical stimulators and associated devices	E011903	Syncope/Fainting	A071302	Inappropriate/Inadequate Shock/Stimulation		
							F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A071302	Inappropriate/Inadequate Shock/Stimulation		
							F26	No Health Consequences or Impact	A071302	Inappropriate/Inadequate Shock/Stimulation		
					CT2866	Implantable cardiac electrical stimulators	E011903	Syncope/Fainting	A071302	Inappropriate/Inadequate Shock/Stimulation		
							F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A071302	Inappropriate/Inadequate Shock/Stimulation		
							F26	No Health Consequences or Impact	A071302	Inappropriate/Inadequate Shock/Stimulation		
					CT2985	Active	E011903	Syncope/Fainting	A071302	Inappropriate/Inadequate Shock/Stimulation		
							F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A071302	Inappropriate/Inadequate Shock/Stimulation		
							F26	No Health Consequences or Impact	A071302	Inappropriate/Inadequate Shock/Stimulation		
					CT3001	Implantable metal	E011903	Syncope/Fainting	A071302	Inappropriate/Inadequate Shock/Stimulation		
							F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A071302	Inappropriate/Inadequate Shock/Stimulation		
							F26	No Health Consequences or Impact	A071302	Inappropriate/Inadequate Shock/Stimulation		
					CT3002	Implantable synthetic polymer	E011903	Syncope/Fainting	A071302	Inappropriate/Inadequate Shock/Stimulation		
							F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A071302	Inappropriate/Inadequate Shock/Stimulation		
							F26	No Health Consequences or Impact	A071302	Inappropriate/Inadequate Shock/Stimulation		
					CT3006	Implantable inorganic material	E011903	Syncope/Fainting	A071302	Inappropriate/Inadequate Shock/Stimulation		
							F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A071302	Inappropriate/Inadequate Shock/Stimulation		
							F26	No Health Consequences or Impact	A071302	Inappropriate/Inadequate Shock/Stimulation		
					CT314	Electrically powered	E011903	Syncope/Fainting	A071302	Inappropriate/Inadequate Shock/Stimulation		
							F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A071302	Inappropriate/Inadequate Shock/Stimulation		
							F26	No Health Consequences or Impact	A071302	Inappropriate/Inadequate Shock/Stimulation		
					CT321	Long-term surgical invasive	E011903	Syncope/Fainting	A071302	Inappropriate/Inadequate Shock/Stimulation		
							F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A071302	Inappropriate/Inadequate Shock/Stimulation		
							F26	No Health Consequences or Impact	A071302	Inappropriate/Inadequate Shock/Stimulation		
					CT449	Generators and associated devices	E011903	Syncope/Fainting	A071302	Inappropriate/Inadequate Shock/Stimulation		
							F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A071302	Inappropriate/Inadequate Shock/Stimulation		
							F26	No Health Consequences or Impact	A071302	Inappropriate/Inadequate Shock/Stimulation		
					CT981	Single use	E011903	Syncope/Fainting	A071302	Inappropriate/Inadequate Shock/Stimulation		
							F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A071302	Inappropriate/Inadequate Shock/Stimulation		
							F26	No Health Consequences or Impact	A071302	Inappropriate/Inadequate Shock/Stimulation		
					CT983	Surgical invasive	E011903	Syncope/Fainting	A071302	Inappropriate/Inadequate Shock/Stimulation		
							F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A071302	Inappropriate/Inadequate Shock/Stimulation		
							F26	No Health Consequences or Impact	A071302	Inappropriate/Inadequate Shock/Stimulation		
10773	2021	All other reportable incidents	35974	Cardiac electrophysiology stimulation system	CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A0712	Pacing Problem		
							F26	No Health Consequences or Impact	A0712	Pacing Problem		
10774	2021	All other reportable incidents	35223	Endocardial/interventricular septal pacing lead	CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A0701	Capturing Problem	05/03/2021	05/03/2021
									A072202	Low Impedance	05/03/2021	05/03/2021

							F10	Inadequate/inappropriate Treatment or Diagnostic Exposure	A0701	Capturing Problem	05/03/2021	05/03/2021
									A072202	Low Impedance	05/03/2021	05/03/2021
							F1901	Additional Surgery	A0701	Capturing Problem	05/03/2021	05/03/2021
									A072202	Low Impedance	05/03/2021	05/03/2021
							F1905	Device Revision or Replacement	A0701	Capturing Problem	05/03/2021	05/03/2021
									A072202	Low Impedance	05/03/2021	05/03/2021
10775	2021	All other reportable incidents	47265	Dual-chamber implantable pacemaker, rate-responsive	CT141	Electrodes/Leads and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	03/03/2021	03/03/2021
							F1901	Additional Surgery	A070504	Premature Discharge of Battery	03/03/2021	03/03/2021
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	03/03/2021	03/03/2021
					CT145	Cardiovascular devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	03/03/2021	03/03/2021
							F1901	Additional Surgery	A070504	Premature Discharge of Battery	03/03/2021	03/03/2021
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	03/03/2021	03/03/2021
					CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	03/03/2021	03/03/2021
							F1901	Additional Surgery	A070504	Premature Discharge of Battery	03/03/2021	03/03/2021
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	03/03/2021	03/03/2021
					CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	03/03/2021	03/03/2021
							F1901	Additional Surgery	A070504	Premature Discharge of Battery	03/03/2021	03/03/2021
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	03/03/2021	03/03/2021
					CT233	Surgical	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	03/03/2021	03/03/2021
							F1901	Additional Surgery	A070504	Premature Discharge of Battery	03/03/2021	03/03/2021
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	03/03/2021	03/03/2021
					CT2406	Implantable	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	03/03/2021	03/03/2021
							F1901	Additional Surgery	A070504	Premature Discharge of Battery	03/03/2021	03/03/2021
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	03/03/2021	03/03/2021
					CT2407	Active implantable	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	03/03/2021	03/03/2021
							F1901	Additional Surgery	A070504	Premature Discharge of Battery	03/03/2021	03/03/2021
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	03/03/2021	03/03/2021
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	03/03/2021	03/03/2021
							F1901	Additional Surgery	A070504	Premature Discharge of Battery	03/03/2021	03/03/2021
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	03/03/2021	03/03/2021
					CT2727	Active implant devices/ancillaries	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	03/03/2021	03/03/2021
							F1901	Additional Surgery	A070504	Premature Discharge of Battery	03/03/2021	03/03/2021
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	03/03/2021	03/03/2021
					CT2750	Implantable cardiac electrical stimulators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	03/03/2021	03/03/2021
							F1901	Additional Surgery	A070504	Premature Discharge of Battery	03/03/2021	03/03/2021
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	03/03/2021	03/03/2021
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	03/03/2021	03/03/2021
							F1901	Additional Surgery	A070504	Premature Discharge of Battery	03/03/2021	03/03/2021
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	03/03/2021	03/03/2021
					CT2985	Active	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	03/03/2021	03/03/2021
							F1901	Additional Surgery	A070504	Premature Discharge of Battery	03/03/2021	03/03/2021
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	03/03/2021	03/03/2021
					CT3001	Implantable metal	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	03/03/2021	03/03/2021
							F1901	Additional Surgery	A070504	Premature Discharge of Battery	03/03/2021	03/03/2021
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	03/03/2021	03/03/2021
					CT3002	Implantable synthetic polymer	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	03/03/2021	03/03/2021
							F1901	Additional Surgery	A070504	Premature Discharge of Battery	03/03/2021	03/03/2021
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	03/03/2021	03/03/2021
					CT3006	Implantable inorganic material	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	03/03/2021	03/03/2021
							F1901	Additional Surgery	A070504	Premature Discharge of Battery	03/03/2021	03/03/2021
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	03/03/2021	03/03/2021
					CT314	Electrically powered	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	03/03/2021	03/03/2021
							F1901	Additional Surgery	A070504	Premature Discharge of Battery	03/03/2021	03/03/2021
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	03/03/2021	03/03/2021
					CT321	Long-term surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	03/03/2021	03/03/2021
							F1901	Additional Surgery	A070504	Premature Discharge of Battery	03/03/2021	03/03/2021
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	03/03/2021	03/03/2021
					CT449	Generators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	03/03/2021	03/03/2021
							F1901	Additional Surgery	A070504	Premature Discharge of Battery	03/03/2021	03/03/2021
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	03/03/2021	03/03/2021
					CT981	Single use	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	03/03/2021	03/03/2021
							F1901	Additional Surgery	A070504	Premature Discharge of Battery	03/03/2021	03/03/2021
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	03/03/2021	03/03/2021
					CT983	Surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	03/03/2021	03/03/2021
							F1901	Additional Surgery	A070504	Premature Discharge of Battery	03/03/2021	03/03/2021
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	03/03/2021	03/03/2021
10776	2021	All other reportable incidents	47263	Cardiac resynchronization therapy implantable pacemaker	CT1684	Cardiac pacemakers and associated devices	E011903	Syncope/Fainting	A071102	Failure to Interrogate	18/02/2021	18/02/2021
									A071208	Pacing Intermittently	18/02/2021	18/02/2021
							F1901	Additional Surgery	A071102	Failure to Interrogate	18/02/2021	18/02/2021
									A071208	Pacing Intermittently	18/02/2021	18/02/2021
							F1905	Device Revision or Replacement	A071102	Failure to Interrogate	18/02/2021	18/02/2021
									A071208	Pacing Intermittently	18/02/2021	18/02/2021
					CT1685	Cardiac pacemakers	E011903	Syncope/Fainting	A071102	Failure to Interrogate	18/02/2021	18/02/2021
									A071208	Pacing Intermittently	18/02/2021	18/02/2021
							F1901	Additional Surgery	A071102	Failure to Interrogate	18/02/2021	18/02/2021
									A071208	Pacing Intermittently	18/02/2021	18/02/2021
							F1905	Device Revision or Replacement	A071102	Failure to Interrogate	18/02/2021	18/02/2021
									A071208	Pacing Intermittently	18/02/2021	18/02/2021
					CT257	Implantable cardiac pacemakers	E011903	Syncope/Fainting	A071102	Failure to Interrogate	18/02/2021	18/02/2021
									A071208	Pacing Intermittently	18/02/2021	18/02/2021
							F1901	Additional Surgery	A071102	Failure to Interrogate	18/02/2021	18/02/2021
									A071208	Pacing Intermittently	18/02/2021	18/02/2021
							F1905	Device Revision or Replacement	A071102	Failure to Interrogate	18/02/2021	18/02/2021
									A071208	Pacing Intermittently	18/02/2021	18/02/2021
					CT2866	Implantable cardiac electrical stimulators	E011903	Syncope/Fainting	A071102	Failure to Interrogate	18/02/2021	18/02/2021
									A071208	Pacing Intermittently	18/02/2021	18/02/2021
							F1901	Additional Surgery	A071102	Failure to Interrogate	18/02/2021	18/02/2021
									A071208	Pacing Intermittently	18/02/2021	18/02/2021

							F1905	Device Revision or Replacement	A071102	Failure to Interrogate	18/02/2021	18/02/2021
									A071208	Pacing Intermittently	18/02/2021	18/02/2021
10777	2021	All other reportable incidents	17882	Physiologic-monitoring defibrillation system	CT1684	Cardiac pacemakers and associated devices			A0721	Circuit Failure		
							E2403	No Clinical Signs, Symptoms or Conditions	A0721	Circuit Failure		
							F27	Problem Identified During Non-Clinical Procedure	A0721	Circuit Failure		
10778	2021	All other reportable incidents	17882	Physiologic-monitoring defibrillation system	CT1684	Cardiac pacemakers and associated devices			A0721	Circuit Failure		
							E2403	No Clinical Signs, Symptoms or Conditions	A0721	Circuit Failure		
10779	2021	All other reportable incidents	17882	Physiologic-monitoring defibrillation system	CT1684	Cardiac pacemakers and associated devices			A1205	Loose or Intermittent Connection		
							E2403	No Clinical Signs, Symptoms or Conditions	A1205	Loose or Intermittent Connection		
10780	2021	All other reportable incidents	17882	Physiologic-monitoring defibrillation system	CT1684	Cardiac pacemakers and associated devices			A1205	Loose or Intermittent Connection		
							E2403	No Clinical Signs, Symptoms or Conditions	A1205	Loose or Intermittent Connection		
10781	2021	All other reportable incidents	17882	Physiologic-monitoring defibrillation system	CT1684	Cardiac pacemakers and associated devices			A1205	Loose or Intermittent Connection		
							E2403	No Clinical Signs, Symptoms or Conditions	A1205	Loose or Intermittent Connection		
10782	2021	All other reportable incidents	35852	Single-chamber implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A070909	Oversensing		
							F26	No Health Consequences or Impact	A070909	Oversensing		
10783	2021	All other reportable incidents	47265	Dual-chamber implantable pacemaker, rate-responsive	CT141	Electrodes/Leads and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem	29/04/2021	29/04/2021
									A070504	Premature Discharge of Battery	29/04/2021	29/04/2021
							F08	Hospitalization or Prolonged Hospitalization	A0705	Battery Problem	29/04/2021	29/04/2021
									A070504	Premature Discharge of Battery	29/04/2021	29/04/2021
							F1905	Device Revision or Replacement	A0705	Battery Problem	29/04/2021	29/04/2021
									A070504	Premature Discharge of Battery	29/04/2021	29/04/2021
					CT145	Cardiovascular devices	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem	29/04/2021	29/04/2021
									A070504	Premature Discharge of Battery	29/04/2021	29/04/2021
							F08	Hospitalization or Prolonged Hospitalization	A0705	Battery Problem	29/04/2021	29/04/2021
									A070504	Premature Discharge of Battery	29/04/2021	29/04/2021
							F1905	Device Revision or Replacement	A0705	Battery Problem	29/04/2021	29/04/2021
									A070504	Premature Discharge of Battery	29/04/2021	29/04/2021
					CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem	29/04/2021	29/04/2021
									A070504	Premature Discharge of Battery	29/04/2021	29/04/2021
							F08	Hospitalization or Prolonged Hospitalization	A0705	Battery Problem	29/04/2021	29/04/2021
									A070504	Premature Discharge of Battery	29/04/2021	29/04/2021
							F1905	Device Revision or Replacement	A0705	Battery Problem	29/04/2021	29/04/2021
									A070504	Premature Discharge of Battery	29/04/2021	29/04/2021
					CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem	29/04/2021	29/04/2021
									A070504	Premature Discharge of Battery	29/04/2021	29/04/2021
							F08	Hospitalization or Prolonged Hospitalization	A0705	Battery Problem	29/04/2021	29/04/2021
									A070504	Premature Discharge of Battery	29/04/2021	29/04/2021
							F1905	Device Revision or Replacement	A0705	Battery Problem	29/04/2021	29/04/2021
									A070504	Premature Discharge of Battery	29/04/2021	29/04/2021
					CT233	Surgical	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem	29/04/2021	29/04/2021
									A070504	Premature Discharge of Battery	29/04/2021	29/04/2021
							F08	Hospitalization or Prolonged Hospitalization	A0705	Battery Problem	29/04/2021	29/04/2021
									A070504	Premature Discharge of Battery	29/04/2021	29/04/2021
							F1905	Device Revision or Replacement	A0705	Battery Problem	29/04/2021	29/04/2021
									A070504	Premature Discharge of Battery	29/04/2021	29/04/2021
					CT2406	Implantable	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem	29/04/2021	29/04/2021
									A070504	Premature Discharge of Battery	29/04/2021	29/04/2021
							F08	Hospitalization or Prolonged Hospitalization	A0705	Battery Problem	29/04/2021	29/04/2021
									A070504	Premature Discharge of Battery	29/04/2021	29/04/2021
							F1905	Device Revision or Replacement	A0705	Battery Problem	29/04/2021	29/04/2021
									A070504	Premature Discharge of Battery	29/04/2021	29/04/2021
					CT2407	Active implantable	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem	29/04/2021	29/04/2021
									A070504	Premature Discharge of Battery	29/04/2021	29/04/2021
							F08	Hospitalization or Prolonged Hospitalization	A0705	Battery Problem	29/04/2021	29/04/2021
									A070504	Premature Discharge of Battery	29/04/2021	29/04/2021
							F1905	Device Revision or Replacement	A0705	Battery Problem	29/04/2021	29/04/2021
									A070504	Premature Discharge of Battery	29/04/2021	29/04/2021
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem	29/04/2021	29/04/2021
									A070504	Premature Discharge of Battery	29/04/2021	29/04/2021
							F08	Hospitalization or Prolonged Hospitalization	A0705	Battery Problem	29/04/2021	29/04/2021
									A070504	Premature Discharge of Battery	29/04/2021	29/04/2021
							F1905	Device Revision or Replacement	A0705	Battery Problem	29/04/2021	29/04/2021
									A070504	Premature Discharge of Battery	29/04/2021	29/04/2021
					CT2727	Active implant devices/ancillaries	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem	29/04/2021	29/04/2021
									A070504	Premature Discharge of Battery	29/04/2021	29/04/2021
							F08	Hospitalization or Prolonged Hospitalization	A0705	Battery Problem	29/04/2021	29/04/2021
									A070504	Premature Discharge of Battery	29/04/2021	29/04/2021
							F1905	Device Revision or Replacement	A0705	Battery Problem	29/04/2021	29/04/2021
									A070504	Premature Discharge of Battery	29/04/2021	29/04/2021
					CT2750	Implantable cardiac electrical stimulators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem	29/04/2021	29/04/2021
									A070504	Premature Discharge of Battery	29/04/2021	29/04/2021
							F08	Hospitalization or Prolonged Hospitalization	A0705	Battery Problem	29/04/2021	29/04/2021
									A070504	Premature Discharge of Battery	29/04/2021	29/04/2021
							F1905	Device Revision or Replacement	A0705	Battery Problem	29/04/2021	29/04/2021
									A070504	Premature Discharge of Battery	29/04/2021	29/04/2021
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem	29/04/2021	29/04/2021
									A070504	Premature Discharge of Battery	29/04/2021	29/04/2021
							F08	Hospitalization or Prolonged Hospitalization	A0705	Battery Problem	29/04/2021	29/04/2021
									A070504	Premature Discharge of Battery	29/04/2021	29/04/2021
							F1905	Device Revision or Replacement	A0705	Battery Problem	29/04/2021	29/04/2021
									A070504	Premature Discharge of Battery	29/04/2021	29/04/2021
					CT2985	Active	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem	29/04/2021	29/04/2021
									A070504	Premature Discharge of Battery	29/04/2021	29/04/2021
							F08	Hospitalization or Prolonged Hospitalization	A0705	Battery Problem	29/04/2021	29/04/2021
									A070504	Premature Discharge of Battery	29/04/2021	29/04/2021
							F1905	Device Revision or Replacement	A0705	Battery Problem	29/04/2021	29/04/2021
									A070504	Premature Discharge of Battery	29/04/2021	29/04/2021
					CT3001	Implantable metal	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem	29/04/2021	29/04/2021

									A070504	Premature Discharge of Battery	29/04/2021	29/04/2021
							F08	Hospitalization or Prolonged Hospitalization	A0705	Battery Problem	29/04/2021	29/04/2021
									A070504	Premature Discharge of Battery	29/04/2021	29/04/2021
							F1905	Device Revision or Replacement	A0705	Battery Problem	29/04/2021	29/04/2021
									A070504	Premature Discharge of Battery	29/04/2021	29/04/2021
					CT3002	Implantable synthetic polymer	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem	29/04/2021	29/04/2021
									A070504	Premature Discharge of Battery	29/04/2021	29/04/2021
							F08	Hospitalization or Prolonged Hospitalization	A0705	Battery Problem	29/04/2021	29/04/2021
									A070504	Premature Discharge of Battery	29/04/2021	29/04/2021
							F1905	Device Revision or Replacement	A0705	Battery Problem	29/04/2021	29/04/2021
									A070504	Premature Discharge of Battery	29/04/2021	29/04/2021
					CT3006	Implantable inorganic material	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem	29/04/2021	29/04/2021
									A070504	Premature Discharge of Battery	29/04/2021	29/04/2021
							F08	Hospitalization or Prolonged Hospitalization	A0705	Battery Problem	29/04/2021	29/04/2021
									A070504	Premature Discharge of Battery	29/04/2021	29/04/2021
							F1905	Device Revision or Replacement	A0705	Battery Problem	29/04/2021	29/04/2021
									A070504	Premature Discharge of Battery	29/04/2021	29/04/2021
					CT314	Electrically powered	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem	29/04/2021	29/04/2021
									A070504	Premature Discharge of Battery	29/04/2021	29/04/2021
							F08	Hospitalization or Prolonged Hospitalization	A0705	Battery Problem	29/04/2021	29/04/2021
									A070504	Premature Discharge of Battery	29/04/2021	29/04/2021
							F1905	Device Revision or Replacement	A0705	Battery Problem	29/04/2021	29/04/2021
									A070504	Premature Discharge of Battery	29/04/2021	29/04/2021
					CT321	Long-term surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem	29/04/2021	29/04/2021
									A070504	Premature Discharge of Battery	29/04/2021	29/04/2021
							F08	Hospitalization or Prolonged Hospitalization	A0705	Battery Problem	29/04/2021	29/04/2021
									A070504	Premature Discharge of Battery	29/04/2021	29/04/2021
							F1905	Device Revision or Replacement	A0705	Battery Problem	29/04/2021	29/04/2021
									A070504	Premature Discharge of Battery	29/04/2021	29/04/2021
					CT449	Generators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem	29/04/2021	29/04/2021
									A070504	Premature Discharge of Battery	29/04/2021	29/04/2021
							F08	Hospitalization or Prolonged Hospitalization	A0705	Battery Problem	29/04/2021	29/04/2021
									A070504	Premature Discharge of Battery	29/04/2021	29/04/2021
							F1905	Device Revision or Replacement	A0705	Battery Problem	29/04/2021	29/04/2021
									A070504	Premature Discharge of Battery	29/04/2021	29/04/2021
					CT981	Single use	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem	29/04/2021	29/04/2021
									A070504	Premature Discharge of Battery	29/04/2021	29/04/2021
							F08	Hospitalization or Prolonged Hospitalization	A0705	Battery Problem	29/04/2021	29/04/2021
									A070504	Premature Discharge of Battery	29/04/2021	29/04/2021
							F1905	Device Revision or Replacement	A0705	Battery Problem	29/04/2021	29/04/2021
									A070504	Premature Discharge of Battery	29/04/2021	29/04/2021
					CT983	Surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem	29/04/2021	29/04/2021
									A070504	Premature Discharge of Battery	29/04/2021	29/04/2021
							F08	Hospitalization or Prolonged Hospitalization	A0705	Battery Problem	29/04/2021	29/04/2021
									A070504	Premature Discharge of Battery	29/04/2021	29/04/2021
							F1905	Device Revision or Replacement	A0705	Battery Problem	29/04/2021	29/04/2021
									A070504	Premature Discharge of Battery	29/04/2021	29/04/2021
10784	2021	All other reportable incidents	47265	Dual-chamber implantable pacemaker, rate-responsive	CT141	Electrodes/Leads and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	08/07/2021	08/07/2021
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	08/07/2021	08/07/2021
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	08/07/2021	08/07/2021
					CT145	Cardiovascular devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	08/07/2021	08/07/2021
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	08/07/2021	08/07/2021
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	08/07/2021	08/07/2021
					CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	08/07/2021	08/07/2021
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	08/07/2021	08/07/2021
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	08/07/2021	08/07/2021
					CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	08/07/2021	08/07/2021
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	08/07/2021	08/07/2021
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	08/07/2021	08/07/2021
					CT233	Surgical	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	08/07/2021	08/07/2021
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	08/07/2021	08/07/2021
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	08/07/2021	08/07/2021
					CT2406	Implantable	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	08/07/2021	08/07/2021
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	08/07/2021	08/07/2021
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	08/07/2021	08/07/2021
					CT2407	Active implantable	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	08/07/2021	08/07/2021
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	08/07/2021	08/07/2021
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	08/07/2021	08/07/2021
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	08/07/2021	08/07/2021
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	08/07/2021	08/07/2021
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	08/07/2021	08/07/2021
					CT2727	Active implant devices/ancillaries	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	08/07/2021	08/07/2021
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	08/07/2021	08/07/2021
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	08/07/2021	08/07/2021
					CT2750	Implantable cardiac electrical stimulators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	08/07/2021	08/07/2021
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	08/07/2021	08/07/2021
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	08/07/2021	08/07/2021
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	08/07/2021	08/07/2021
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	08/07/2021	08/07/2021
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	08/07/2021	08/07/2021
					CT2985	Active	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	08/07/2021	08/07/2021
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	08/07/2021	08/07/2021
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	08/07/2021	08/07/2021
					CT3001	Implantable metal	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	08/07/2021	08/07/2021
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	08/07/2021	08/07/2021
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	08/07/2021	08/07/2021
					CT3002	Implantable synthetic polymer	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	08/07/2021	08/07/2021
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	08/07/2021	08/07/2021

						F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	08/07/2021	08/07/2021
				CT3006	Implantable inorganic material	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	08/07/2021	08/07/2021
						F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	08/07/2021	08/07/2021
						F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	08/07/2021	08/07/2021
				CT314	Electrically powered	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	08/07/2021	08/07/2021
						F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	08/07/2021	08/07/2021
						F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	08/07/2021	08/07/2021
				CT321	Long-term surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	08/07/2021	08/07/2021
						F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	08/07/2021	08/07/2021
						F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	08/07/2021	08/07/2021
				CT449	Generators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	08/07/2021	08/07/2021
						F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	08/07/2021	08/07/2021
						F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	08/07/2021	08/07/2021
				CT981	Single use	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	08/07/2021	08/07/2021
						F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	08/07/2021	08/07/2021
						F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	08/07/2021	08/07/2021
				CT983	Surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	08/07/2021	08/07/2021
						F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	08/07/2021	08/07/2021
						F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	08/07/2021	08/07/2021
10785	2021	All other reportable incidents	36241	Pacemaker/defibrillator lead	CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A040101	Fracture	30/04/2021	30/04/2021
									A072201	High Impedance	30/04/2021	30/04/2021
							F23	Unexpected Medical Intervention	A040101	Fracture	30/04/2021	30/04/2021
									A072201	High Impedance	30/04/2021	30/04/2021
10786	2021	All other reportable incidents	47270	Cardiac resynchronization therapy implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A070909	Oversensing		
									A0712	Pacing Problem		
							F26	No Health Consequences or Impact	A070909	Oversensing		
									A0712	Pacing Problem		
10787	2021	All other reportable incidents	47263	Cardiac resynchronization therapy implantable pacemaker	CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070909	Oversensing		
									A0712	Pacing Problem		
									A2304	Off-Label Use		
							F26	No Health Consequences or Impact	A070909	Oversensing		
									A0712	Pacing Problem		
									A2304	Off-Label Use		
				CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A070909	Oversensing		
									A0712	Pacing Problem		
									A2304	Off-Label Use		
							F26	No Health Consequences or Impact	A070909	Oversensing		
									A0712	Pacing Problem		
									A2304	Off-Label Use		
				CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A070909	Oversensing		
									A0712	Pacing Problem		
									A2304	Off-Label Use		
							F26	No Health Consequences or Impact	A070909	Oversensing		
									A0712	Pacing Problem		
									A2304	Off-Label Use		
				CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A070909	Oversensing		
									A0712	Pacing Problem		
									A2304	Off-Label Use		
							F26	No Health Consequences or Impact	A070909	Oversensing		
									A0712	Pacing Problem		
									A2304	Off-Label Use		
10788	2021	All other reportable incidents	47265	Dual-chamber implantable pacemaker, rate-responsive	CT141	Electrodes/Leads and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT145	Cardiovascular devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT233	Surgical	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT2406	Implantable	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT2407	Active implantable	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT2727	Active implant devices/ancillaries	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT2750	Implantable cardiac electrical stimulators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT2985	Active	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT3001	Implantable metal	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT3002	Implantable synthetic polymer	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT3006	Implantable inorganic material	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT314	Electrically powered	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT321	Long-term surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT449	Generators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT981	Single use	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		

							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT983	Surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
10789	2021	All other reportable incidents	35223	Endocardial/interventricular septal pacing lead	CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070909	Oversensing	10/03/2021	10/03/2021
							F1901	Additional Surgery	A070909	Oversensing	10/03/2021	10/03/2021
							F1905	Device Revision or Replacement	A070909	Oversensing	10/03/2021	10/03/2021
10790	2021	All other reportable incidents	47265	Dual-chamber implantable pacemaker, rate-responsive	CT141	Electrodes/Leads and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A160601	Premature Elective Replacement Indicator		
							F1903	Device Explantation	A160601	Premature Elective Replacement Indicator		
					CT145	Cardiovascular devices	E2403	No Clinical Signs, Symptoms or Conditions	A160601	Premature Elective Replacement Indicator		
							F1903	Device Explantation	A160601	Premature Elective Replacement Indicator		
					CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A160601	Premature Elective Replacement Indicator		
							F1903	Device Explantation	A160601	Premature Elective Replacement Indicator		
					CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A160601	Premature Elective Replacement Indicator		
							F1903	Device Explantation	A160601	Premature Elective Replacement Indicator		
					CT233	Surgical	E2403	No Clinical Signs, Symptoms or Conditions	A160601	Premature Elective Replacement Indicator		
							F1903	Device Explantation	A160601	Premature Elective Replacement Indicator		
					CT2406	Implantable	E2403	No Clinical Signs, Symptoms or Conditions	A160601	Premature Elective Replacement Indicator		
							F1903	Device Explantation	A160601	Premature Elective Replacement Indicator		
					CT2407	Active implantable	E2403	No Clinical Signs, Symptoms or Conditions	A160601	Premature Elective Replacement Indicator		
							F1903	Device Explantation	A160601	Premature Elective Replacement Indicator		
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A160601	Premature Elective Replacement Indicator		
							F1903	Device Explantation	A160601	Premature Elective Replacement Indicator		
					CT2727	Active implant devices/ancillaries	E2403	No Clinical Signs, Symptoms or Conditions	A160601	Premature Elective Replacement Indicator		
							F1903	Device Explantation	A160601	Premature Elective Replacement Indicator		
					CT2750	Implantable cardiac electrical stimulators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A160601	Premature Elective Replacement Indicator		
							F1903	Device Explantation	A160601	Premature Elective Replacement Indicator		
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A160601	Premature Elective Replacement Indicator		
							F1903	Device Explantation	A160601	Premature Elective Replacement Indicator		
					CT2985	Active	E2403	No Clinical Signs, Symptoms or Conditions	A160601	Premature Elective Replacement Indicator		
							F1903	Device Explantation	A160601	Premature Elective Replacement Indicator		
					CT3001	Implantable metal	E2403	No Clinical Signs, Symptoms or Conditions	A160601	Premature Elective Replacement Indicator		
							F1903	Device Explantation	A160601	Premature Elective Replacement Indicator		
					CT3002	Implantable synthetic polymer	E2403	No Clinical Signs, Symptoms or Conditions	A160601	Premature Elective Replacement Indicator		
							F1903	Device Explantation	A160601	Premature Elective Replacement Indicator		
					CT3006	Implantable inorganic material	E2403	No Clinical Signs, Symptoms or Conditions	A160601	Premature Elective Replacement Indicator		
							F1903	Device Explantation	A160601	Premature Elective Replacement Indicator		
					CT314	Electrically powered	E2403	No Clinical Signs, Symptoms or Conditions	A160601	Premature Elective Replacement Indicator		
							F1903	Device Explantation	A160601	Premature Elective Replacement Indicator		
					CT321	Long-term surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A160601	Premature Elective Replacement Indicator		
							F1903	Device Explantation	A160601	Premature Elective Replacement Indicator		
					CT449	Generators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A160601	Premature Elective Replacement Indicator		
							F1903	Device Explantation	A160601	Premature Elective Replacement Indicator		
					CT981	Single use	E2403	No Clinical Signs, Symptoms or Conditions	A160601	Premature Elective Replacement Indicator		
							F1903	Device Explantation	A160601	Premature Elective Replacement Indicator		
					CT983	Surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A160601	Premature Elective Replacement Indicator		
							F1903	Device Explantation	A160601	Premature Elective Replacement Indicator		
10791	2021	All other reportable incidents	17882	Physiologic-monitoring defibrillation system	CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070908	Failure to Sense		
							F26	No Health Consequences or Impact	A070908	Failure to Sense		
10792	2021	All other reportable incidents	37265	Dual-chamber implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	18/03/2021	18/03/2021
							F1901	Additional Surgery	A070504	Premature Discharge of Battery	18/03/2021	18/03/2021
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	18/03/2021	18/03/2021
10793	2021	All other reportable incidents	36241	Pacemaker/defibrillator lead	CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A0203	Defective Device	12/03/2021	12/03/2021
									A070102	High Capture Threshold	12/03/2021	12/03/2021
									A072201	High Impedance	12/03/2021	12/03/2021
									A090801	Signal Artifact/Noise	12/03/2021	12/03/2021
									A1301	Failure to Read Input Signal	12/03/2021	12/03/2021
							F1908	Prolonged Surgery	A0203	Defective Device	12/03/2021	12/03/2021
									A070102	High Capture Threshold	12/03/2021	12/03/2021
									A072201	High Impedance	12/03/2021	12/03/2021
									A090801	Signal Artifact/Noise	12/03/2021	12/03/2021
									A1301	Failure to Read Input Signal	12/03/2021	12/03/2021
10794	2021	All other reportable incidents	36241	Pacemaker/defibrillator lead	CT1684	Cardiac pacemakers and associated devices	E0602	Cardiac Arrest	A070909	Oversensing		
							E2403	No Clinical Signs, Symptoms or Conditions	A070909	Oversensing		
							F26	No Health Consequences or Impact	A070909	Oversensing		
10795	2021	All other reportable incidents	35852	Single-chamber implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem	16/04/2021	16/04/2021
									A070504	Premature Discharge of Battery	16/04/2021	16/04/2021
							F08	Hospitalization or Prolonged Hospitalization	A0705	Battery Problem	16/04/2021	16/04/2021
									A070504	Premature Discharge of Battery	16/04/2021	16/04/2021
							F12	Serious Injury/ Illness/ Impairment	A0705	Battery Problem	16/04/2021	16/04/2021
									A070504	Premature Discharge of Battery	16/04/2021	16/04/2021
							F19	Surgical Intervention	A0705	Battery Problem	16/04/2021	16/04/2021
									A070504	Premature Discharge of Battery	16/04/2021	16/04/2021
10796	2021	All other reportable incidents	36241	Pacemaker/defibrillator lead	CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A072201	High Impedance		
							F26	No Health Consequences or Impact	A072201	High Impedance		
10797	2021	All other reportable incidents	17882	Physiologic-monitoring defibrillation system	CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A07	Electrical /Electronic Property Problem		
							F27	Problem Identified During Non-Clinical Procedure	A07	Electrical /Electronic Property Problem		
10798	2021	All other reportable incidents	35223	Endocardial/interventricular septal pacing lead	CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A07	Electrical /Electronic Property Problem		
									A072202	Low Impedance		
							F26	No Health Consequences or Impact	A07	Electrical /Electronic Property Problem		
									A072202	Low Impedance		
10799	2021	All other reportable incidents	36241	Pacemaker/defibrillator lead	CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A05	Mechanical Problem	14/01/2020	14/01/2020
							F1905	Device Revision or Replacement	A05	Mechanical Problem	14/01/2020	14/01/2020
							F1908	Prolonged Surgery	A05	Mechanical Problem	14/01/2020	14/01/2020
10800	2021	All other reportable incidents	47265	Dual-chamber implantable pacemaker, rate-responsive	CT141	Electrodes/Leads and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
							F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A0705	Battery Problem		

								A070504	Premature Discharge of Battery		
					CT145	Cardiovascular devices	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
							F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
					CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
							F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
					CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
							F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
					CT233	Surgical	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
							F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
					CT2406	Implantable	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
							F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
					CT2407	Active implantable	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
							F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
							F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
					CT2727	Active implant devices/ancillaries	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
							F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
					CT2750	Implantable cardiac electrical stimulators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
							F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
							F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
					CT2985	Active	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
							F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
					CT3001	Implantable metal	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
							F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
					CT3002	Implantable synthetic polymer	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
							F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
					CT3006	Implantable inorganic material	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
							F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
					CT314	Electrically powered	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
							F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A0705	Battery Problem		

									A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
					CT321	Long-term surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
							F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
					CT449	Generators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
							F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
					CT981	Single use	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
							F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
					CT983	Surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
							F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
10801	2021	All other reportable incidents	37265	Dual-chamber implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	24/05/2021	24/05/2021
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	24/05/2021	24/05/2021
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	24/05/2021	24/05/2021
10802	2021	All other reportable incidents	35852	Single-chamber implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	01/04/2021	01/04/2021
							F19	Surgical Intervention	A070504	Premature Discharge of Battery	01/04/2021	01/04/2021
10803	2021	All other reportable incidents	35223	Endocardial/interventricular septal pacing lead	CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A090801	Signal Artifact/Noise		
							F26	No Health Consequences or Impact	A090801	Signal Artifact/Noise		
10804	2021	All other reportable incidents	47265	Dual-chamber implantable pacemaker, rate-responsive	CT141	Electrodes/Leads and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	03/05/2021	03/05/2021
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	03/05/2021	03/05/2021
							F12	Serious Injury/ Illness/ Impairment	A070504	Premature Discharge of Battery	03/05/2021	03/05/2021
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	03/05/2021	03/05/2021
					CT145	Cardiovascular devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	03/05/2021	03/05/2021
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	03/05/2021	03/05/2021
							F12	Serious Injury/ Illness/ Impairment	A070504	Premature Discharge of Battery	03/05/2021	03/05/2021
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	03/05/2021	03/05/2021
					CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	03/05/2021	03/05/2021
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	03/05/2021	03/05/2021
							F12	Serious Injury/ Illness/ Impairment	A070504	Premature Discharge of Battery	03/05/2021	03/05/2021
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	03/05/2021	03/05/2021
					CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	03/05/2021	03/05/2021
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	03/05/2021	03/05/2021
							F12	Serious Injury/ Illness/ Impairment	A070504	Premature Discharge of Battery	03/05/2021	03/05/2021
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	03/05/2021	03/05/2021
					CT233	Surgical	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	03/05/2021	03/05/2021
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	03/05/2021	03/05/2021
							F12	Serious Injury/ Illness/ Impairment	A070504	Premature Discharge of Battery	03/05/2021	03/05/2021
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	03/05/2021	03/05/2021
					CT2406	Implantable	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	03/05/2021	03/05/2021
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	03/05/2021	03/05/2021
							F12	Serious Injury/ Illness/ Impairment	A070504	Premature Discharge of Battery	03/05/2021	03/05/2021
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	03/05/2021	03/05/2021
					CT2407	Active implantable	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	03/05/2021	03/05/2021
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	03/05/2021	03/05/2021
							F12	Serious Injury/ Illness/ Impairment	A070504	Premature Discharge of Battery	03/05/2021	03/05/2021
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	03/05/2021	03/05/2021
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	03/05/2021	03/05/2021
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	03/05/2021	03/05/2021
							F12	Serious Injury/ Illness/ Impairment	A070504	Premature Discharge of Battery	03/05/2021	03/05/2021
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	03/05/2021	03/05/2021
					CT2727	Active implant devices/ancillaries	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	03/05/2021	03/05/2021
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	03/05/2021	03/05/2021
							F12	Serious Injury/ Illness/ Impairment	A070504	Premature Discharge of Battery	03/05/2021	03/05/2021
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	03/05/2021	03/05/2021
					CT2750	Implantable cardiac electrical stimulators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	03/05/2021	03/05/2021
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	03/05/2021	03/05/2021
							F12	Serious Injury/ Illness/ Impairment	A070504	Premature Discharge of Battery	03/05/2021	03/05/2021
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	03/05/2021	03/05/2021
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	03/05/2021	03/05/2021
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	03/05/2021	03/05/2021
							F12	Serious Injury/ Illness/ Impairment	A070504	Premature Discharge of Battery	03/05/2021	03/05/2021
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	03/05/2021	03/05/2021
					CT2985	Active	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	03/05/2021	03/05/2021
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	03/05/2021	03/05/2021
							F12	Serious Injury/ Illness/ Impairment	A070504	Premature Discharge of Battery	03/05/2021	03/05/2021
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	03/05/2021	03/05/2021
					CT3001	Implantable metal	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	03/05/2021	03/05/2021
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	03/05/2021	03/05/2021
							F12	Serious Injury/ Illness/ Impairment	A070504	Premature Discharge of Battery	03/05/2021	03/05/2021
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	03/05/2021	03/05/2021
					CT3002	Implantable synthetic polymer	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	03/05/2021	03/05/2021
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	03/05/2021	03/05/2021

						F12	Serious Injury/ Illness/ Impairment	A070504	Premature Discharge of Battery	03/05/2021	03/05/2021
						F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	03/05/2021	03/05/2021
					CT3006	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	03/05/2021	03/05/2021
						F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	03/05/2021	03/05/2021
						F12	Serious Injury/ Illness/ Impairment	A070504	Premature Discharge of Battery	03/05/2021	03/05/2021
						F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	03/05/2021	03/05/2021
					CT314	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	03/05/2021	03/05/2021
						F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	03/05/2021	03/05/2021
						F12	Serious Injury/ Illness/ Impairment	A070504	Premature Discharge of Battery	03/05/2021	03/05/2021
						F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	03/05/2021	03/05/2021
					CT321	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	03/05/2021	03/05/2021
						F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	03/05/2021	03/05/2021
						F12	Serious Injury/ Illness/ Impairment	A070504	Premature Discharge of Battery	03/05/2021	03/05/2021
						F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	03/05/2021	03/05/2021
					CT449	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	03/05/2021	03/05/2021
						F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	03/05/2021	03/05/2021
						F12	Serious Injury/ Illness/ Impairment	A070504	Premature Discharge of Battery	03/05/2021	03/05/2021
						F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	03/05/2021	03/05/2021
					CT981	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	03/05/2021	03/05/2021
						F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	03/05/2021	03/05/2021
						F12	Serious Injury/ Illness/ Impairment	A070504	Premature Discharge of Battery	03/05/2021	03/05/2021
						F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	03/05/2021	03/05/2021
					CT983	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	03/05/2021	03/05/2021
						F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	03/05/2021	03/05/2021
						F12	Serious Injury/ Illness/ Impairment	A070504	Premature Discharge of Battery	03/05/2021	03/05/2021
						F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	03/05/2021	03/05/2021
10805	2021	All other reportable incidents	47267	Single-chamber implantable pacemaker, rate-responsive	CT141	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	05/07/2021	05/07/2021
						F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	05/07/2021	05/07/2021
						F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	05/07/2021	05/07/2021
					CT145	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	05/07/2021	05/07/2021
						F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	05/07/2021	05/07/2021
						F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	05/07/2021	05/07/2021
					CT1684	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	05/07/2021	05/07/2021
						F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	05/07/2021	05/07/2021
						F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	05/07/2021	05/07/2021
					CT1685	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	05/07/2021	05/07/2021
						F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	05/07/2021	05/07/2021
						F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	05/07/2021	05/07/2021
					CT233	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	05/07/2021	05/07/2021
						F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	05/07/2021	05/07/2021
						F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	05/07/2021	05/07/2021
					CT2406	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	05/07/2021	05/07/2021
						F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	05/07/2021	05/07/2021
						F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	05/07/2021	05/07/2021
					CT2407	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	05/07/2021	05/07/2021
						F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	05/07/2021	05/07/2021
						F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	05/07/2021	05/07/2021
					CT257	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	05/07/2021	05/07/2021
						F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	05/07/2021	05/07/2021
						F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	05/07/2021	05/07/2021
					CT2727	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	05/07/2021	05/07/2021
						F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	05/07/2021	05/07/2021
						F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	05/07/2021	05/07/2021
					CT2750	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	05/07/2021	05/07/2021
						F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	05/07/2021	05/07/2021
						F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	05/07/2021	05/07/2021
					CT2866	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	05/07/2021	05/07/2021
						F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	05/07/2021	05/07/2021
						F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	05/07/2021	05/07/2021
					CT2985	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	05/07/2021	05/07/2021
						F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	05/07/2021	05/07/2021
						F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	05/07/2021	05/07/2021
					CT3001	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	05/07/2021	05/07/2021
						F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	05/07/2021	05/07/2021
						F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	05/07/2021	05/07/2021
					CT3002	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	05/07/2021	05/07/2021
						F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	05/07/2021	05/07/2021
						F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	05/07/2021	05/07/2021
					CT3006	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	05/07/2021	05/07/2021
						F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	05/07/2021	05/07/2021
						F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	05/07/2021	05/07/2021
					CT314	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	05/07/2021	05/07/2021
						F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	05/07/2021	05/07/2021
						F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	05/07/2021	05/07/2021
					CT321	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	05/07/2021	05/07/2021
						F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	05/07/2021	05/07/2021
						F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	05/07/2021	05/07/2021
					CT449	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	05/07/2021	05/07/2021
						F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	05/07/2021	05/07/2021
						F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	05/07/2021	05/07/2021
					CT981	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	05/07/2021	05/07/2021
						F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	05/07/2021	05/07/2021
						F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	05/07/2021	05/07/2021
					CT983	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	05/07/2021	05/07/2021
						F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	05/07/2021	05/07/2021
						F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	05/07/2021	05/07/2021
10806	2021	All other reportable incidents	47267	Single-chamber implantable pacemaker, rate-responsive	CT141	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	25/05/2021	25/05/2021
						F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	25/05/2021	25/05/2021

							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	25/05/2021	25/05/2021
					CT145	Cardiovascular devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	25/05/2021	25/05/2021
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	25/05/2021	25/05/2021
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	25/05/2021	25/05/2021
					CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	25/05/2021	25/05/2021
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	25/05/2021	25/05/2021
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	25/05/2021	25/05/2021
					CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	25/05/2021	25/05/2021
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	25/05/2021	25/05/2021
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	25/05/2021	25/05/2021
					CT233	Surgical	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	25/05/2021	25/05/2021
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	25/05/2021	25/05/2021
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	25/05/2021	25/05/2021
					CT2406	Implantable	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	25/05/2021	25/05/2021
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	25/05/2021	25/05/2021
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	25/05/2021	25/05/2021
					CT2407	Active implantable	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	25/05/2021	25/05/2021
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	25/05/2021	25/05/2021
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	25/05/2021	25/05/2021
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	25/05/2021	25/05/2021
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	25/05/2021	25/05/2021
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	25/05/2021	25/05/2021
					CT2727	Active implant devices/ancillaries	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	25/05/2021	25/05/2021
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	25/05/2021	25/05/2021
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	25/05/2021	25/05/2021
					CT2750	Implantable cardiac electrical stimulators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	25/05/2021	25/05/2021
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	25/05/2021	25/05/2021
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	25/05/2021	25/05/2021
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	25/05/2021	25/05/2021
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	25/05/2021	25/05/2021
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	25/05/2021	25/05/2021
					CT2985	Active	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	25/05/2021	25/05/2021
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	25/05/2021	25/05/2021
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	25/05/2021	25/05/2021
					CT3001	Implantable metal	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	25/05/2021	25/05/2021
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	25/05/2021	25/05/2021
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	25/05/2021	25/05/2021
					CT3002	Implantable synthetic polymer	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	25/05/2021	25/05/2021
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	25/05/2021	25/05/2021
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	25/05/2021	25/05/2021
					CT3006	Implantable inorganic material	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	25/05/2021	25/05/2021
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	25/05/2021	25/05/2021
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	25/05/2021	25/05/2021
					CT314	Electrically powered	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	25/05/2021	25/05/2021
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	25/05/2021	25/05/2021
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	25/05/2021	25/05/2021
					CT321	Long-term surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	25/05/2021	25/05/2021
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	25/05/2021	25/05/2021
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	25/05/2021	25/05/2021
					CT449	Generators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	25/05/2021	25/05/2021
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	25/05/2021	25/05/2021
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	25/05/2021	25/05/2021
					CT981	Single use	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	25/05/2021	25/05/2021
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	25/05/2021	25/05/2021
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	25/05/2021	25/05/2021
					CT983	Surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	25/05/2021	25/05/2021
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	25/05/2021	25/05/2021
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	25/05/2021	25/05/2021
10807	2021	All other reportable incidents	35852	Single-chamber implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
									A070602	Delayed Charge Time		
									A0908	Incorrect, Inadequate or Imprecise Result or Readings		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
									A070602	Delayed Charge Time		
									A0908	Incorrect, Inadequate or Imprecise Result or Readings		
10808	2021	All other reportable incidents	47265	Dual-chamber implantable pacemaker, rate-responsive	CT141	Electrodes/Leads and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	16/04/2021	16/04/2021
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	16/04/2021	16/04/2021
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	16/04/2021	16/04/2021
					CT145	Cardiovascular devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	16/04/2021	16/04/2021
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	16/04/2021	16/04/2021
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	16/04/2021	16/04/2021
					CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	16/04/2021	16/04/2021
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	16/04/2021	16/04/2021
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	16/04/2021	16/04/2021
					CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	16/04/2021	16/04/2021
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	16/04/2021	16/04/2021
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	16/04/2021	16/04/2021
					CT233	Surgical	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	16/04/2021	16/04/2021
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	16/04/2021	16/04/2021
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	16/04/2021	16/04/2021
					CT2406	Implantable	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	16/04/2021	16/04/2021
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	16/04/2021	16/04/2021
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	16/04/2021	16/04/2021
					CT2407	Active implantable	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	16/04/2021	16/04/2021
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	16/04/2021	16/04/2021
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	16/04/2021	16/04/2021
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	16/04/2021	16/04/2021

							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	16/04/2021	16/04/2021
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	16/04/2021	16/04/2021
					CT2727	Active implant devices/ancillaries	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	16/04/2021	16/04/2021
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	16/04/2021	16/04/2021
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	16/04/2021	16/04/2021
					CT2750	Implantable cardiac electrical stimulators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	16/04/2021	16/04/2021
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	16/04/2021	16/04/2021
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	16/04/2021	16/04/2021
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	16/04/2021	16/04/2021
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	16/04/2021	16/04/2021
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	16/04/2021	16/04/2021
					CT2985	Active	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	16/04/2021	16/04/2021
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	16/04/2021	16/04/2021
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	16/04/2021	16/04/2021
					CT3001	Implantable metal	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	16/04/2021	16/04/2021
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	16/04/2021	16/04/2021
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	16/04/2021	16/04/2021
					CT3002	Implantable synthetic polymer	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	16/04/2021	16/04/2021
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	16/04/2021	16/04/2021
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	16/04/2021	16/04/2021
					CT3006	Implantable inorganic material	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	16/04/2021	16/04/2021
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	16/04/2021	16/04/2021
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	16/04/2021	16/04/2021
					CT314	Electrically powered	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	16/04/2021	16/04/2021
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	16/04/2021	16/04/2021
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	16/04/2021	16/04/2021
					CT321	Long-term surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	16/04/2021	16/04/2021
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	16/04/2021	16/04/2021
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	16/04/2021	16/04/2021
					CT449	Generators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	16/04/2021	16/04/2021
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	16/04/2021	16/04/2021
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	16/04/2021	16/04/2021
					CT981	Single use	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	16/04/2021	16/04/2021
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	16/04/2021	16/04/2021
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	16/04/2021	16/04/2021
					CT983	Surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	16/04/2021	16/04/2021
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	16/04/2021	16/04/2021
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	16/04/2021	16/04/2021
10809	2021	All other reportable incidents	35852	Single-chamber implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	18/03/2021	18/03/2021
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	18/03/2021	18/03/2021
							F19	Surgical Intervention	A070504	Premature Discharge of Battery	18/03/2021	18/03/2021
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	18/03/2021	18/03/2021
10810	2021	All other reportable incidents	47265	Dual-chamber implantable pacemaker, rate-responsive	CT141	Electrodes/Leads and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem	07/04/2021	07/04/2021
									A070504	Premature Discharge of Battery	07/04/2021	07/04/2021
							F08	Hospitalization or Prolonged Hospitalization	A0705	Battery Problem	07/04/2021	07/04/2021
									A070504	Premature Discharge of Battery	07/04/2021	07/04/2021
							F1903	Device Explantation	A0705	Battery Problem	07/04/2021	07/04/2021
									A070504	Premature Discharge of Battery	07/04/2021	07/04/2021
					CT145	Cardiovascular devices	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem	07/04/2021	07/04/2021
									A070504	Premature Discharge of Battery	07/04/2021	07/04/2021
							F08	Hospitalization or Prolonged Hospitalization	A0705	Battery Problem	07/04/2021	07/04/2021
									A070504	Premature Discharge of Battery	07/04/2021	07/04/2021
							F1903	Device Explantation	A0705	Battery Problem	07/04/2021	07/04/2021
									A070504	Premature Discharge of Battery	07/04/2021	07/04/2021
					CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem	07/04/2021	07/04/2021
									A070504	Premature Discharge of Battery	07/04/2021	07/04/2021
							F08	Hospitalization or Prolonged Hospitalization	A0705	Battery Problem	07/04/2021	07/04/2021
									A070504	Premature Discharge of Battery	07/04/2021	07/04/2021
							F1903	Device Explantation	A0705	Battery Problem	07/04/2021	07/04/2021
									A070504	Premature Discharge of Battery	07/04/2021	07/04/2021
					CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem	07/04/2021	07/04/2021
									A070504	Premature Discharge of Battery	07/04/2021	07/04/2021
							F08	Hospitalization or Prolonged Hospitalization	A0705	Battery Problem	07/04/2021	07/04/2021
									A070504	Premature Discharge of Battery	07/04/2021	07/04/2021
							F1903	Device Explantation	A0705	Battery Problem	07/04/2021	07/04/2021
									A070504	Premature Discharge of Battery	07/04/2021	07/04/2021
					CT233	Surgical	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem	07/04/2021	07/04/2021
									A070504	Premature Discharge of Battery	07/04/2021	07/04/2021
							F08	Hospitalization or Prolonged Hospitalization	A0705	Battery Problem	07/04/2021	07/04/2021
									A070504	Premature Discharge of Battery	07/04/2021	07/04/2021
							F1903	Device Explantation	A0705	Battery Problem	07/04/2021	07/04/2021
									A070504	Premature Discharge of Battery	07/04/2021	07/04/2021
					CT2406	Implantable	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem	07/04/2021	07/04/2021
									A070504	Premature Discharge of Battery	07/04/2021	07/04/2021
							F08	Hospitalization or Prolonged Hospitalization	A0705	Battery Problem	07/04/2021	07/04/2021
									A070504	Premature Discharge of Battery	07/04/2021	07/04/2021
							F1903	Device Explantation	A0705	Battery Problem	07/04/2021	07/04/2021
									A070504	Premature Discharge of Battery	07/04/2021	07/04/2021
					CT2407	Active implantable	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem	07/04/2021	07/04/2021
									A070504	Premature Discharge of Battery	07/04/2021	07/04/2021
							F08	Hospitalization or Prolonged Hospitalization	A0705	Battery Problem	07/04/2021	07/04/2021
									A070504	Premature Discharge of Battery	07/04/2021	07/04/2021
							F1903	Device Explantation	A0705	Battery Problem	07/04/2021	07/04/2021
									A070504	Premature Discharge of Battery	07/04/2021	07/04/2021
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem	07/04/2021	07/04/2021
									A070504	Premature Discharge of Battery	07/04/2021	07/04/2021
							F08	Hospitalization or Prolonged Hospitalization	A0705	Battery Problem	07/04/2021	07/04/2021
									A070504	Premature Discharge of Battery	07/04/2021	07/04/2021

							F1903	Device Explantation	A0705	Battery Problem	07/04/2021	07/04/2021
									A070504	Premature Discharge of Battery	07/04/2021	07/04/2021
					CT2727	Active implant devices/ancillaries	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem	07/04/2021	07/04/2021
									A070504	Premature Discharge of Battery	07/04/2021	07/04/2021
							F08	Hospitalization or Prolonged Hospitalization	A0705	Battery Problem	07/04/2021	07/04/2021
									A070504	Premature Discharge of Battery	07/04/2021	07/04/2021
							F1903	Device Explantation	A0705	Battery Problem	07/04/2021	07/04/2021
									A070504	Premature Discharge of Battery	07/04/2021	07/04/2021
					CT2750	Implantable cardiac electrical stimulators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem	07/04/2021	07/04/2021
									A070504	Premature Discharge of Battery	07/04/2021	07/04/2021
							F08	Hospitalization or Prolonged Hospitalization	A0705	Battery Problem	07/04/2021	07/04/2021
									A070504	Premature Discharge of Battery	07/04/2021	07/04/2021
							F1903	Device Explantation	A0705	Battery Problem	07/04/2021	07/04/2021
									A070504	Premature Discharge of Battery	07/04/2021	07/04/2021
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem	07/04/2021	07/04/2021
									A070504	Premature Discharge of Battery	07/04/2021	07/04/2021
							F08	Hospitalization or Prolonged Hospitalization	A0705	Battery Problem	07/04/2021	07/04/2021
									A070504	Premature Discharge of Battery	07/04/2021	07/04/2021
							F1903	Device Explantation	A0705	Battery Problem	07/04/2021	07/04/2021
									A070504	Premature Discharge of Battery	07/04/2021	07/04/2021
					CT2985	Active	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem	07/04/2021	07/04/2021
									A070504	Premature Discharge of Battery	07/04/2021	07/04/2021
							F08	Hospitalization or Prolonged Hospitalization	A0705	Battery Problem	07/04/2021	07/04/2021
									A070504	Premature Discharge of Battery	07/04/2021	07/04/2021
							F1903	Device Explantation	A0705	Battery Problem	07/04/2021	07/04/2021
									A070504	Premature Discharge of Battery	07/04/2021	07/04/2021
					CT3001	Implantable metal	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem	07/04/2021	07/04/2021
									A070504	Premature Discharge of Battery	07/04/2021	07/04/2021
							F08	Hospitalization or Prolonged Hospitalization	A0705	Battery Problem	07/04/2021	07/04/2021
									A070504	Premature Discharge of Battery	07/04/2021	07/04/2021
							F1903	Device Explantation	A0705	Battery Problem	07/04/2021	07/04/2021
									A070504	Premature Discharge of Battery	07/04/2021	07/04/2021
					CT3002	Implantable synthetic polymer	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem	07/04/2021	07/04/2021
									A070504	Premature Discharge of Battery	07/04/2021	07/04/2021
							F08	Hospitalization or Prolonged Hospitalization	A0705	Battery Problem	07/04/2021	07/04/2021
									A070504	Premature Discharge of Battery	07/04/2021	07/04/2021
							F1903	Device Explantation	A0705	Battery Problem	07/04/2021	07/04/2021
									A070504	Premature Discharge of Battery	07/04/2021	07/04/2021
					CT3006	Implantable inorganic material	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem	07/04/2021	07/04/2021
									A070504	Premature Discharge of Battery	07/04/2021	07/04/2021
							F08	Hospitalization or Prolonged Hospitalization	A0705	Battery Problem	07/04/2021	07/04/2021
									A070504	Premature Discharge of Battery	07/04/2021	07/04/2021
							F1903	Device Explantation	A0705	Battery Problem	07/04/2021	07/04/2021
									A070504	Premature Discharge of Battery	07/04/2021	07/04/2021
					CT314	Electrically powered	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem	07/04/2021	07/04/2021
									A070504	Premature Discharge of Battery	07/04/2021	07/04/2021
							F08	Hospitalization or Prolonged Hospitalization	A0705	Battery Problem	07/04/2021	07/04/2021
									A070504	Premature Discharge of Battery	07/04/2021	07/04/2021
							F1903	Device Explantation	A0705	Battery Problem	07/04/2021	07/04/2021
									A070504	Premature Discharge of Battery	07/04/2021	07/04/2021
					CT321	Long-term surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem	07/04/2021	07/04/2021
									A070504	Premature Discharge of Battery	07/04/2021	07/04/2021
							F08	Hospitalization or Prolonged Hospitalization	A0705	Battery Problem	07/04/2021	07/04/2021
									A070504	Premature Discharge of Battery	07/04/2021	07/04/2021
							F1903	Device Explantation	A0705	Battery Problem	07/04/2021	07/04/2021
									A070504	Premature Discharge of Battery	07/04/2021	07/04/2021
					CT449	Generators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem	07/04/2021	07/04/2021
									A070504	Premature Discharge of Battery	07/04/2021	07/04/2021
							F08	Hospitalization or Prolonged Hospitalization	A0705	Battery Problem	07/04/2021	07/04/2021
									A070504	Premature Discharge of Battery	07/04/2021	07/04/2021
							F1903	Device Explantation	A0705	Battery Problem	07/04/2021	07/04/2021
									A070504	Premature Discharge of Battery	07/04/2021	07/04/2021
					CT981	Single use	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem	07/04/2021	07/04/2021
									A070504	Premature Discharge of Battery	07/04/2021	07/04/2021
							F08	Hospitalization or Prolonged Hospitalization	A0705	Battery Problem	07/04/2021	07/04/2021
									A070504	Premature Discharge of Battery	07/04/2021	07/04/2021
							F1903	Device Explantation	A0705	Battery Problem	07/04/2021	07/04/2021
									A070504	Premature Discharge of Battery	07/04/2021	07/04/2021
					CT983	Surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem	07/04/2021	07/04/2021
									A070504	Premature Discharge of Battery	07/04/2021	07/04/2021
							F08	Hospitalization or Prolonged Hospitalization	A0705	Battery Problem	07/04/2021	07/04/2021
									A070504	Premature Discharge of Battery	07/04/2021	07/04/2021
							F1903	Device Explantation	A0705	Battery Problem	07/04/2021	07/04/2021
									A070504	Premature Discharge of Battery	07/04/2021	07/04/2021
10811	2021	All other reportable incidents	37265	Dual-chamber implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A070909	Oversensing		
									A090801	Signal Artifact/Noise		
							F26	No Health Consequences or Impact	A070909	Oversensing		
									A090801	Signal Artifact/Noise		
10812	2021	All other reportable incidents	35852	Single-chamber implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E2401	Insufficient Information	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
							E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
							F24	Insufficient Health Impact Information	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
10813	2021	All other reportable incidents	35852	Single-chamber implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	15/01/2021	15/01/2021
							F19	Surgical Intervention	A070504	Premature Discharge of Battery	15/01/2021	15/01/2021
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	15/01/2021	15/01/2021
10814	2021	All other reportable incidents	36241	Pacemaker/defibrillator lead	CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070909	Oversensing		

									A090801	Signal Artifact/Noise		
							F15	Recognized Device or Procedural Complication	A070909	Oversensing		
									A090801	Signal Artifact/Noise		
10815	2021	All other reportable incidents	35223	Endocardial/interventricular septal pacing lead	CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070102	High Capture Threshold		
									A072201	High Impedance		
							F26	No Health Consequences or Impact	A070102	High Capture Threshold		
									A072201	High Impedance		
10816	2021	All other reportable incidents	35223	Endocardial/interventricular septal pacing lead	CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A051101	Collapse		
									A0722	Impedance Problem		
									A072201	High Impedance		
							F1905	Device Revision or Replacement	A051101	Collapse		
									A0722	Impedance Problem		
									A072201	High Impedance		
							F26	No Health Consequences or Impact	A051101	Collapse		
									A0722	Impedance Problem		
									A072201	High Impedance		
10817	2021	All other reportable incidents	63188	Left-ventricular wireless pacing system sensor/transmitter	CT141	Electrodes/Leads and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070801	Complete Loss of Power		
					CT145	Cardiovascular devices	E2403	No Clinical Signs, Symptoms or Conditions	A070801	Complete Loss of Power		
					CT1577	Wireless monitoring system transmitters	E2403	No Clinical Signs, Symptoms or Conditions	A070801	Complete Loss of Power		
					CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070801	Complete Loss of Power		
					CT2317	Transmitters/Receivers/Transceivers	E2403	No Clinical Signs, Symptoms or Conditions	A070801	Complete Loss of Power		
					CT233	Surgical	E2403	No Clinical Signs, Symptoms or Conditions	A070801	Complete Loss of Power		
					CT2406	Implantable	E2403	No Clinical Signs, Symptoms or Conditions	A070801	Complete Loss of Power		
					CT2407	Active implantable	E2403	No Clinical Signs, Symptoms or Conditions	A070801	Complete Loss of Power		
					CT255	Transducers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070801	Complete Loss of Power		
					CT2726	Physiological monitoring/measuring devices	E2403	No Clinical Signs, Symptoms or Conditions	A070801	Complete Loss of Power		
					CT2727	Active implant devices/ancillaries	E2403	No Clinical Signs, Symptoms or Conditions	A070801	Complete Loss of Power		
					CT2750	Implantable cardiac electrical stimulators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070801	Complete Loss of Power		
					CT2885	Physiological monitoring/measuring system transmitters/receivers/transceivers	E2403	No Clinical Signs, Symptoms or Conditions	A070801	Complete Loss of Power		
					CT2985	Active	E2403	No Clinical Signs, Symptoms or Conditions	A070801	Complete Loss of Power		
					CT3001	Implantable metal	E2403	No Clinical Signs, Symptoms or Conditions	A070801	Complete Loss of Power		
					CT3002	Implantable synthetic polymer	E2403	No Clinical Signs, Symptoms or Conditions	A070801	Complete Loss of Power		
					CT3006	Implantable inorganic material	E2403	No Clinical Signs, Symptoms or Conditions	A070801	Complete Loss of Power		
					CT314	Electrically powered	E2403	No Clinical Signs, Symptoms or Conditions	A070801	Complete Loss of Power		
					CT321	Long-term surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A070801	Complete Loss of Power		
					CT384	Sensors and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070801	Complete Loss of Power		
					CT449	Generators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070801	Complete Loss of Power		
					CT981	Single use	E2403	No Clinical Signs, Symptoms or Conditions	A070801	Complete Loss of Power		
					CT983	Surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A070801	Complete Loss of Power		
					CT994	Sensors	E2403	No Clinical Signs, Symptoms or Conditions	A070801	Complete Loss of Power		
10818	2021	All other reportable incidents	47270	Cardiac resynchronization therapy implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E2401	Insufficient Information	A070909	Oversensing		
									A072201	High Impedance		
									A090801	Signal Artifact/Noise		
							E2403	No Clinical Signs, Symptoms or Conditions	A070909	Oversensing		
									A072201	High Impedance		
									A090801	Signal Artifact/Noise		
							F26	No Health Consequences or Impact	A070909	Oversensing		
									A072201	High Impedance		
									A090801	Signal Artifact/Noise		
10819	2021	All other reportable incidents	36241	Pacemaker/defibrillator lead	CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A05	Mechanical Problem	22/10/2020	22/10/2020
							F1908	Prolonged Surgery	A05	Mechanical Problem	22/10/2020	22/10/2020
10820	2021	All other reportable incidents	47265	Dual-chamber implantable pacemaker, rate-responsive	CT141	Electrodes/Leads and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070909	Oversensing	16/03/2021	16/03/2021
							F1901	Additional Surgery	A070909	Oversensing	16/03/2021	16/03/2021
							F1905	Device Revision or Replacement	A070909	Oversensing	16/03/2021	16/03/2021
					CT145	Cardiovascular devices	E2403	No Clinical Signs, Symptoms or Conditions	A070909	Oversensing	16/03/2021	16/03/2021
							F1901	Additional Surgery	A070909	Oversensing	16/03/2021	16/03/2021
							F1905	Device Revision or Replacement	A070909	Oversensing	16/03/2021	16/03/2021
					CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070909	Oversensing	16/03/2021	16/03/2021
							F1901	Additional Surgery	A070909	Oversensing	16/03/2021	16/03/2021
							F1905	Device Revision or Replacement	A070909	Oversensing	16/03/2021	16/03/2021
					CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A070909	Oversensing	16/03/2021	16/03/2021
							F1901	Additional Surgery	A070909	Oversensing	16/03/2021	16/03/2021
							F1905	Device Revision or Replacement	A070909	Oversensing	16/03/2021	16/03/2021
					CT233	Surgical	E2403	No Clinical Signs, Symptoms or Conditions	A070909	Oversensing	16/03/2021	16/03/2021
							F1901	Additional Surgery	A070909	Oversensing	16/03/2021	16/03/2021
							F1905	Device Revision or Replacement	A070909	Oversensing	16/03/2021	16/03/2021
					CT2406	Implantable	E2403	No Clinical Signs, Symptoms or Conditions	A070909	Oversensing	16/03/2021	16/03/2021
							F1901	Additional Surgery	A070909	Oversensing	16/03/2021	16/03/2021
							F1905	Device Revision or Replacement	A070909	Oversensing	16/03/2021	16/03/2021
					CT2407	Active implantable	E2403	No Clinical Signs, Symptoms or Conditions	A070909	Oversensing	16/03/2021	16/03/2021
							F1901	Additional Surgery	A070909	Oversensing	16/03/2021	16/03/2021
							F1905	Device Revision or Replacement	A070909	Oversensing	16/03/2021	16/03/2021
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A070909	Oversensing	16/03/2021	16/03/2021
							F1901	Additional Surgery	A070909	Oversensing	16/03/2021	16/03/2021
							F1905	Device Revision or Replacement	A070909	Oversensing	16/03/2021	16/03/2021
					CT2727	Active implant devices/ancillaries	E2403	No Clinical Signs, Symptoms or Conditions	A070909	Oversensing	16/03/2021	16/03/2021
							F1901	Additional Surgery	A070909	Oversensing	16/03/2021	16/03/2021
							F1905	Device Revision or Replacement	A070909	Oversensing	16/03/2021	16/03/2021
					CT2750	Implantable cardiac electrical stimulators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070909	Oversensing	16/03/2021	16/03/2021
							F1901	Additional Surgery	A070909	Oversensing	16/03/2021	16/03/2021
							F1905	Device Revision or Replacement	A070909	Oversensing	16/03/2021	16/03/2021
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A070909	Oversensing	16/03/2021	16/03/2021
							F1901	Additional Surgery	A070909	Oversensing	16/03/2021	16/03/2021
							F1905	Device Revision or Replacement	A070909	Oversensing	16/03/2021	16/03/2021
					CT2985	Active	E2403	No Clinical Signs, Symptoms or Conditions	A070909	Oversensing	16/03/2021	16/03/2021
							F1901	Additional Surgery	A070909	Oversensing	16/03/2021	16/03/2021

							F1905	Device Revision or Replacement	A070909	Oversensing	16/03/2021	16/03/2021
					CT3001	Implantable metal	E2403	No Clinical Signs, Symptoms or Conditions	A070909	Oversensing	16/03/2021	16/03/2021
							F1901	Additional Surgery	A070909	Oversensing	16/03/2021	16/03/2021
							F1905	Device Revision or Replacement	A070909	Oversensing	16/03/2021	16/03/2021
					CT3002	Implantable synthetic polymer	E2403	No Clinical Signs, Symptoms or Conditions	A070909	Oversensing	16/03/2021	16/03/2021
							F1901	Additional Surgery	A070909	Oversensing	16/03/2021	16/03/2021
							F1905	Device Revision or Replacement	A070909	Oversensing	16/03/2021	16/03/2021
					CT3006	Implantable inorganic material	E2403	No Clinical Signs, Symptoms or Conditions	A070909	Oversensing	16/03/2021	16/03/2021
							F1901	Additional Surgery	A070909	Oversensing	16/03/2021	16/03/2021
							F1905	Device Revision or Replacement	A070909	Oversensing	16/03/2021	16/03/2021
					CT314	Electrically powered	E2403	No Clinical Signs, Symptoms or Conditions	A070909	Oversensing	16/03/2021	16/03/2021
							F1901	Additional Surgery	A070909	Oversensing	16/03/2021	16/03/2021
							F1905	Device Revision or Replacement	A070909	Oversensing	16/03/2021	16/03/2021
					CT321	Long-term surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A070909	Oversensing	16/03/2021	16/03/2021
							F1901	Additional Surgery	A070909	Oversensing	16/03/2021	16/03/2021
							F1905	Device Revision or Replacement	A070909	Oversensing	16/03/2021	16/03/2021
					CT449	Generators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070909	Oversensing	16/03/2021	16/03/2021
							F1901	Additional Surgery	A070909	Oversensing	16/03/2021	16/03/2021
							F1905	Device Revision or Replacement	A070909	Oversensing	16/03/2021	16/03/2021
					CT981	Single use	E2403	No Clinical Signs, Symptoms or Conditions	A070909	Oversensing	16/03/2021	16/03/2021
							F1901	Additional Surgery	A070909	Oversensing	16/03/2021	16/03/2021
							F1905	Device Revision or Replacement	A070909	Oversensing	16/03/2021	16/03/2021
					CT983	Surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A070909	Oversensing	16/03/2021	16/03/2021
							F1901	Additional Surgery	A070909	Oversensing	16/03/2021	16/03/2021
							F1905	Device Revision or Replacement	A070909	Oversensing	16/03/2021	16/03/2021
10821	2021	All other reportable incidents	35223	Endocardial/interventricular septal pacing lead	CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A0401	Break		
									A070909	Oversensing		
							F1901	Additional Surgery	A0401	Break		
									A070909	Oversensing		
10822	2021	All other reportable incidents	35223	Endocardial/interventricular septal pacing lead	CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A0401	Break		
									A070909	Oversensing		
							F1901	Additional Surgery	A0401	Break		
									A070909	Oversensing		
10823	2021	All other reportable incidents	37265	Dual-chamber implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A0706	Charging Problem		
									A070909	Oversensing		
							F22	Unexpected Diagnostic Intervention	A0706	Charging Problem		
									A070909	Oversensing		
10824	2021	All other reportable incidents	17882	Physiologic-monitoring defibrillation system	CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A07	Electrical /Electronic Property Problem		
							F27	Problem Identified During Non-Clinical Procedure	A07	Electrical /Electronic Property Problem		
10825	2021	All other reportable incidents	37265	Dual-chamber implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	19/03/2021	19/03/2021
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	19/03/2021	19/03/2021
							F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A070504	Premature Discharge of Battery	19/03/2021	19/03/2021
							F12	Serious Injury/ Illness/ Impairment	A070504	Premature Discharge of Battery	19/03/2021	19/03/2021
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	19/03/2021	19/03/2021
10826	2021	All other reportable incidents	47265	Dual-chamber implantable pacemaker, rate-responsive	CT141	Electrodes/Leads and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
					CT145	Cardiovascular devices	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
					CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
					CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
					CT233	Surgical	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
					CT2406	Implantable	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
					CT2407	Active implantable	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
					CT2727	Active implant devices/ancillaries	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
					CT2750	Implantable cardiac electrical stimulators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		

					CT2985	Active	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
					CT3001	Implantable metal	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
					CT3002	Implantable synthetic polymer	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
					CT3006	Implantable inorganic material	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
					CT314	Electrically powered	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
					CT321	Long-term surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
					CT449	Generators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
					CT981	Single use	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
					CT983	Surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
10827	2021	All other reportable incidents	47265	Dual-chamber implantable pacemaker, rate-responsive	CT141	Electrodes/Leads and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F27	Problem Identified During Non-Clinical Procedure	A070504	Premature Discharge of Battery		
					CT145	Cardiovascular devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F27	Problem Identified During Non-Clinical Procedure	A070504	Premature Discharge of Battery		
					CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F27	Problem Identified During Non-Clinical Procedure	A070504	Premature Discharge of Battery		
					CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F27	Problem Identified During Non-Clinical Procedure	A070504	Premature Discharge of Battery		
					CT233	Surgical	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F27	Problem Identified During Non-Clinical Procedure	A070504	Premature Discharge of Battery		
					CT2406	Implantable	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F27	Problem Identified During Non-Clinical Procedure	A070504	Premature Discharge of Battery		
					CT2407	Active implantable	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F27	Problem Identified During Non-Clinical Procedure	A070504	Premature Discharge of Battery		
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F27	Problem Identified During Non-Clinical Procedure	A070504	Premature Discharge of Battery		
					CT2727	Active implant devices/ancillaries	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F27	Problem Identified During Non-Clinical Procedure	A070504	Premature Discharge of Battery		
					CT2750	Implantable cardiac electrical stimulators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F27	Problem Identified During Non-Clinical Procedure	A070504	Premature Discharge of Battery		
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F27	Problem Identified During Non-Clinical Procedure	A070504	Premature Discharge of Battery		
					CT2985	Active	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F27	Problem Identified During Non-Clinical Procedure	A070504	Premature Discharge of Battery		
					CT3001	Implantable metal	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F27	Problem Identified During Non-Clinical Procedure	A070504	Premature Discharge of Battery		
					CT3002	Implantable synthetic polymer	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F27	Problem Identified During Non-Clinical Procedure	A070504	Premature Discharge of Battery		
					CT3006	Implantable inorganic material	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F27	Problem Identified During Non-Clinical Procedure	A070504	Premature Discharge of Battery		
					CT314	Electrically powered	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F27	Problem Identified During Non-Clinical Procedure	A070504	Premature Discharge of Battery		
					CT321	Long-term surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F27	Problem Identified During Non-Clinical Procedure	A070504	Premature Discharge of Battery		
					CT449	Generators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F27	Problem Identified During Non-Clinical Procedure	A070504	Premature Discharge of Battery		
					CT981	Single use	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F27	Problem Identified During Non-Clinical Procedure	A070504	Premature Discharge of Battery		
					CT983	Surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F27	Problem Identified During Non-Clinical Procedure	A070504	Premature Discharge of Battery		
10828	2021	All other reportable incidents	35852	Single-chamber implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E2104	Electric Shock	A071302	Inappropriate/Inadequate Shock/Stimulation		
							F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A071302	Inappropriate/Inadequate Shock/Stimulation		
							F26	No Health Consequences or Impact	A071302	Inappropriate/Inadequate Shock/Stimulation		
10829	2021	All other reportable incidents	17882	Physiologic-monitoring defibrillation system	CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A0719	Unexpected Shutdown		
							F26	No Health Consequences or Impact	A0719	Unexpected Shutdown		
10830	2021	All other reportable incidents	35223	Endocardial/interventricular septal pacing lead	CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A040101	Fracture	25/03/2021	25/03/2021
									A07	Electrical /Electronic Property Problem	25/03/2021	25/03/2021
									A070103	Intermittent Capture	25/03/2021	25/03/2021
									A070909	Oversensing	25/03/2021	25/03/2021
									A072201	High Impedance	25/03/2021	25/03/2021
									A090801	Signal Artifact/Noise	25/03/2021	25/03/2021
							F26	No Health Consequences or Impact	A040101	Fracture	25/03/2021	25/03/2021

								A07	Electrical /Electronic Property Problem	25/03/2021	25/03/2021
								A070103	Intermittent Capture	25/03/2021	25/03/2021
								A070909	Oversensing	25/03/2021	25/03/2021
								A072201	High Impedance	25/03/2021	25/03/2021
								A090801	Signal Artifact/Noise	25/03/2021	25/03/2021
10831	2021	All other reportable incidents	35852	Single-chamber implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	22/03/2021	22/03/2021
							F1901	Additional Surgery	A070504	Premature Discharge of Battery	22/03/2021	22/03/2021
10832	2021	All other reportable incidents	47265	Dual-chamber implantable pacemaker, rate-responsive	CT141	Electrodes/Leads and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	23/12/2020	23/12/2020
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery	23/12/2020	23/12/2020
					CT145	Cardiovascular devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	23/12/2020	23/12/2020
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery	23/12/2020	23/12/2020
					CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	23/12/2020	23/12/2020
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery	23/12/2020	23/12/2020
					CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	23/12/2020	23/12/2020
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery	23/12/2020	23/12/2020
					CT233	Surgical	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	23/12/2020	23/12/2020
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery	23/12/2020	23/12/2020
					CT2406	Implantable	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	23/12/2020	23/12/2020
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery	23/12/2020	23/12/2020
					CT2407	Active implantable	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	23/12/2020	23/12/2020
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery	23/12/2020	23/12/2020
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	23/12/2020	23/12/2020
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery	23/12/2020	23/12/2020
					CT2727	Active implant devices/ancillaries	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	23/12/2020	23/12/2020
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery	23/12/2020	23/12/2020
					CT2750	Implantable cardiac electrical stimulators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	23/12/2020	23/12/2020
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery	23/12/2020	23/12/2020
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	23/12/2020	23/12/2020
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery	23/12/2020	23/12/2020
					CT2985	Active	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	23/12/2020	23/12/2020
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery	23/12/2020	23/12/2020
					CT3001	Implantable metal	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	23/12/2020	23/12/2020
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery	23/12/2020	23/12/2020
					CT3002	Implantable synthetic polymer	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	23/12/2020	23/12/2020
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery	23/12/2020	23/12/2020
					CT3006	Implantable inorganic material	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	23/12/2020	23/12/2020
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery	23/12/2020	23/12/2020
					CT314	Electrically powered	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	23/12/2020	23/12/2020
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery	23/12/2020	23/12/2020
					CT321	Long-term surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	23/12/2020	23/12/2020
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery	23/12/2020	23/12/2020
					CT449	Generators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	23/12/2020	23/12/2020
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery	23/12/2020	23/12/2020
					CT981	Single use	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	23/12/2020	23/12/2020
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery	23/12/2020	23/12/2020
					CT983	Surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	23/12/2020	23/12/2020
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery	23/12/2020	23/12/2020
10833	2021	All other reportable incidents	47265	Dual-chamber implantable pacemaker, rate-responsive	CT141	Electrodes/Leads and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT145	Cardiovascular devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT233	Surgical	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT2406	Implantable	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT2407	Active implantable	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT2727	Active implant devices/ancillaries	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT2750	Implantable cardiac electrical stimulators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT2985	Active	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT3001	Implantable metal	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT3002	Implantable synthetic polymer	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT3006	Implantable inorganic material	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT314	Electrically powered	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT321	Long-term surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT449	Generators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT981	Single use	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT983	Surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
10834	2021	All other reportable incidents	35852	Single-chamber implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	09/04/2021	09/04/2021

							F19	Surgical Intervention	A070504	Premature Discharge of Battery	09/04/2021	09/04/2021
10835	2021	All other reportable incidents	47265	Dual-chamber implantable pacemaker, rate-responsive	CT141	Electrodes/Leads and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem		
							F26	No Health Consequences or Impact	A0705	Battery Problem		
					CT145	Cardiovascular devices	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem		
							F26	No Health Consequences or Impact	A0705	Battery Problem		
					CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem		
							F26	No Health Consequences or Impact	A0705	Battery Problem		
					CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem		
							F26	No Health Consequences or Impact	A0705	Battery Problem		
					CT233	Surgical	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem		
							F26	No Health Consequences or Impact	A0705	Battery Problem		
					CT2406	Implantable	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem		
							F26	No Health Consequences or Impact	A0705	Battery Problem		
					CT2407	Active implantable	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem		
							F26	No Health Consequences or Impact	A0705	Battery Problem		
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem		
							F26	No Health Consequences or Impact	A0705	Battery Problem		
					CT2727	Active implant devices/ancillaries	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem		
							F26	No Health Consequences or Impact	A0705	Battery Problem		
					CT2750	Implantable cardiac electrical stimulators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem		
							F26	No Health Consequences or Impact	A0705	Battery Problem		
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem		
							F26	No Health Consequences or Impact	A0705	Battery Problem		
					CT2985	Active	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem		
							F26	No Health Consequences or Impact	A0705	Battery Problem		
					CT3001	Implantable metal	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem		
							F26	No Health Consequences or Impact	A0705	Battery Problem		
					CT3002	Implantable synthetic polymer	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem		
							F26	No Health Consequences or Impact	A0705	Battery Problem		
					CT3006	Implantable inorganic material	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem		
							F26	No Health Consequences or Impact	A0705	Battery Problem		
					CT314	Electrically powered	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem		
							F26	No Health Consequences or Impact	A0705	Battery Problem		
					CT321	Long-term surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem		
							F26	No Health Consequences or Impact	A0705	Battery Problem		
					CT449	Generators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem		
							F26	No Health Consequences or Impact	A0705	Battery Problem		
					CT981	Single use	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem		
							F26	No Health Consequences or Impact	A0705	Battery Problem		
					CT983	Surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem		
							F26	No Health Consequences or Impact	A0705	Battery Problem		
10836	2021	All other reportable incidents	35852	Single-chamber implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem	15/04/2021	15/04/2021
									A070504	Premature Discharge of Battery	15/04/2021	15/04/2021
							F19	Surgical Intervention	A0705	Battery Problem	15/04/2021	15/04/2021
									A070504	Premature Discharge of Battery	15/04/2021	15/04/2021
10837	2021	All other reportable incidents	47265	Dual-chamber implantable pacemaker, rate-responsive	CT141	Electrodes/Leads and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	09/05/2021	09/05/2021
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	09/05/2021	09/05/2021
							F12	Serious Injury/ Illness/ Impairment	A070504	Premature Discharge of Battery	09/05/2021	09/05/2021
							F19	Surgical Intervention	A070504	Premature Discharge of Battery	09/05/2021	09/05/2021
					CT145	Cardiovascular devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	09/05/2021	09/05/2021
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	09/05/2021	09/05/2021
							F12	Serious Injury/ Illness/ Impairment	A070504	Premature Discharge of Battery	09/05/2021	09/05/2021
							F19	Surgical Intervention	A070504	Premature Discharge of Battery	09/05/2021	09/05/2021
					CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	09/05/2021	09/05/2021
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	09/05/2021	09/05/2021
							F12	Serious Injury/ Illness/ Impairment	A070504	Premature Discharge of Battery	09/05/2021	09/05/2021
							F19	Surgical Intervention	A070504	Premature Discharge of Battery	09/05/2021	09/05/2021
					CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	09/05/2021	09/05/2021
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	09/05/2021	09/05/2021
							F12	Serious Injury/ Illness/ Impairment	A070504	Premature Discharge of Battery	09/05/2021	09/05/2021
							F19	Surgical Intervention	A070504	Premature Discharge of Battery	09/05/2021	09/05/2021
					CT233	Surgical	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	09/05/2021	09/05/2021
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	09/05/2021	09/05/2021
							F12	Serious Injury/ Illness/ Impairment	A070504	Premature Discharge of Battery	09/05/2021	09/05/2021
							F19	Surgical Intervention	A070504	Premature Discharge of Battery	09/05/2021	09/05/2021
					CT2406	Implantable	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	09/05/2021	09/05/2021
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	09/05/2021	09/05/2021
							F12	Serious Injury/ Illness/ Impairment	A070504	Premature Discharge of Battery	09/05/2021	09/05/2021
							F19	Surgical Intervention	A070504	Premature Discharge of Battery	09/05/2021	09/05/2021
					CT2407	Active implantable	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	09/05/2021	09/05/2021
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	09/05/2021	09/05/2021
							F12	Serious Injury/ Illness/ Impairment	A070504	Premature Discharge of Battery	09/05/2021	09/05/2021
							F19	Surgical Intervention	A070504	Premature Discharge of Battery	09/05/2021	09/05/2021
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	09/05/2021	09/05/2021
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	09/05/2021	09/05/2021
							F12	Serious Injury/ Illness/ Impairment	A070504	Premature Discharge of Battery	09/05/2021	09/05/2021
							F19	Surgical Intervention	A070504	Premature Discharge of Battery	09/05/2021	09/05/2021
					CT2727	Active implant devices/ancillaries	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	09/05/2021	09/05/2021
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	09/05/2021	09/05/2021
							F12	Serious Injury/ Illness/ Impairment	A070504	Premature Discharge of Battery	09/05/2021	09/05/2021
							F19	Surgical Intervention	A070504	Premature Discharge of Battery	09/05/2021	09/05/2021
					CT2750	Implantable cardiac electrical stimulators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	09/05/2021	09/05/2021
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	09/05/2021	09/05/2021
							F12	Serious Injury/ Illness/ Impairment	A070504	Premature Discharge of Battery	09/05/2021	09/05/2021
							F19	Surgical Intervention	A070504	Premature Discharge of Battery	09/05/2021	09/05/2021
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	09/05/2021	09/05/2021
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	09/05/2021	09/05/2021
							F12	Serious Injury/ Illness/ Impairment	A070504	Premature Discharge of Battery	09/05/2021	09/05/2021

							F19	Surgical Intervention	A070504	Premature Discharge of Battery	09/05/2021	09/05/2021
					CT2985	Active	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	09/05/2021	09/05/2021
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	09/05/2021	09/05/2021
							F12	Serious Injury/ Illness/ Impairment	A070504	Premature Discharge of Battery	09/05/2021	09/05/2021
							F19	Surgical Intervention	A070504	Premature Discharge of Battery	09/05/2021	09/05/2021
					CT3001	Implantable metal	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	09/05/2021	09/05/2021
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	09/05/2021	09/05/2021
							F12	Serious Injury/ Illness/ Impairment	A070504	Premature Discharge of Battery	09/05/2021	09/05/2021
							F19	Surgical Intervention	A070504	Premature Discharge of Battery	09/05/2021	09/05/2021
					CT3002	Implantable synthetic polymer	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	09/05/2021	09/05/2021
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	09/05/2021	09/05/2021
							F12	Serious Injury/ Illness/ Impairment	A070504	Premature Discharge of Battery	09/05/2021	09/05/2021
							F19	Surgical Intervention	A070504	Premature Discharge of Battery	09/05/2021	09/05/2021
					CT3006	Implantable inorganic material	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	09/05/2021	09/05/2021
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	09/05/2021	09/05/2021
							F12	Serious Injury/ Illness/ Impairment	A070504	Premature Discharge of Battery	09/05/2021	09/05/2021
							F19	Surgical Intervention	A070504	Premature Discharge of Battery	09/05/2021	09/05/2021
					CT314	Electrically powered	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	09/05/2021	09/05/2021
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	09/05/2021	09/05/2021
							F12	Serious Injury/ Illness/ Impairment	A070504	Premature Discharge of Battery	09/05/2021	09/05/2021
							F19	Surgical Intervention	A070504	Premature Discharge of Battery	09/05/2021	09/05/2021
					CT321	Long-term surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	09/05/2021	09/05/2021
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	09/05/2021	09/05/2021
							F12	Serious Injury/ Illness/ Impairment	A070504	Premature Discharge of Battery	09/05/2021	09/05/2021
							F19	Surgical Intervention	A070504	Premature Discharge of Battery	09/05/2021	09/05/2021
					CT449	Generators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	09/05/2021	09/05/2021
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	09/05/2021	09/05/2021
							F12	Serious Injury/ Illness/ Impairment	A070504	Premature Discharge of Battery	09/05/2021	09/05/2021
							F19	Surgical Intervention	A070504	Premature Discharge of Battery	09/05/2021	09/05/2021
					CT981	Single use	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	09/05/2021	09/05/2021
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	09/05/2021	09/05/2021
							F12	Serious Injury/ Illness/ Impairment	A070504	Premature Discharge of Battery	09/05/2021	09/05/2021
							F19	Surgical Intervention	A070504	Premature Discharge of Battery	09/05/2021	09/05/2021
					CT983	Surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	09/05/2021	09/05/2021
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	09/05/2021	09/05/2021
							F12	Serious Injury/ Illness/ Impairment	A070504	Premature Discharge of Battery	09/05/2021	09/05/2021
							F19	Surgical Intervention	A070504	Premature Discharge of Battery	09/05/2021	09/05/2021
10838	2021	All other reportable incidents	35852	Single-chamber implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A070909	Oversensing		
							F26	No Health Consequences or Impact	A070909	Oversensing		
10839	2021	All other reportable incidents	35852	Single-chamber implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E2104	Electric Shock	A071303	Intermittent Shock/Stimulation		
							E2403	No Clinical Signs, Symptoms or Conditions	A071303	Intermittent Shock/Stimulation		
							F1905	Device Revision or Replacement	A071303	Intermittent Shock/Stimulation		
10840	2021	All other reportable incidents	47265	Dual-chamber implantable pacemaker, rate-responsive	CT141	Electrodes/Leads and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
					CT145	Cardiovascular devices	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
					CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
					CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
					CT233	Surgical	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
					CT2406	Implantable	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
					CT2407	Active implantable	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
					CT2727	Active implant devices/ancillaries	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
					CT2750	Implantable cardiac electrical stimulators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
					CT2985	Active	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		

							F26	No Health Consequences or Impact	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
					CT3001	Implantable metal	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
					CT3002	Implantable synthetic polymer	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
					CT3006	Implantable inorganic material	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
					CT314	Electrically powered	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
					CT321	Long-term surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
					CT449	Generators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
					CT981	Single use	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
					CT983	Surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
10841	2021	All other reportable incidents	17882	Physiologic-monitoring defibrillation system	CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070803	Failure to Power Up		
							F27	Problem Identified During Non-Clinical Procedure	A070803	Failure to Power Up		
10842	2021	All other reportable incidents	37265	Dual-chamber implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	19/03/2021	19/03/2021
									A071102	Failure to Interrogate	19/03/2021	19/03/2021
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery	19/03/2021	19/03/2021
									A071102	Failure to Interrogate	19/03/2021	19/03/2021
10843	2021	All other reportable incidents	35852	Single-chamber implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E060101	Asystole	A071302	Inappropriate/Inadequate Shock/Stimulation		
							E2104	Electric Shock	A071302	Inappropriate/Inadequate Shock/Stimulation		
							E2403	No Clinical Signs, Symptoms or Conditions	A071302	Inappropriate/Inadequate Shock/Stimulation		
							F26	No Health Consequences or Impact	A071302	Inappropriate/Inadequate Shock/Stimulation		
10844	2021	All other reportable incidents	47265	Dual-chamber implantable pacemaker, rate-responsive	CT141	Electrodes/Leads and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT145	Cardiovascular devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT233	Surgical	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT2406	Implantable	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT2407	Active implantable	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT2727	Active implant devices/ancillaries	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT2750	Implantable cardiac electrical stimulators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT2985	Active	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT3001	Implantable metal	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT3002	Implantable synthetic polymer	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT3006	Implantable inorganic material	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT314	Electrically powered	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT321	Long-term surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT449	Generators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT981	Single use	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT983	Surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
10845	2021	All other reportable incidents	47265	Dual-chamber implantable pacemaker, rate-responsive	CT141	Electrodes/Leads and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	17/12/2020	17/12/2020
							F12	Serious Injury/ Illness/ Impairment	A070504	Premature Discharge of Battery	17/12/2020	17/12/2020
							F19	Surgical Intervention	A070504	Premature Discharge of Battery	17/12/2020	17/12/2020
							F1903	Device Explantation	A070504	Premature Discharge of Battery	17/12/2020	17/12/2020

					CT145	Cardiovascular devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	17/12/2020	17/12/2020
							F12	Serious Injury/ Illness/ Impairment	A070504	Premature Discharge of Battery	17/12/2020	17/12/2020
							F19	Surgical Intervention	A070504	Premature Discharge of Battery	17/12/2020	17/12/2020
							F1903	Device Explantation	A070504	Premature Discharge of Battery	17/12/2020	17/12/2020
					CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	17/12/2020	17/12/2020
							F12	Serious Injury/ Illness/ Impairment	A070504	Premature Discharge of Battery	17/12/2020	17/12/2020
							F19	Surgical Intervention	A070504	Premature Discharge of Battery	17/12/2020	17/12/2020
							F1903	Device Explantation	A070504	Premature Discharge of Battery	17/12/2020	17/12/2020
					CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	17/12/2020	17/12/2020
							F12	Serious Injury/ Illness/ Impairment	A070504	Premature Discharge of Battery	17/12/2020	17/12/2020
							F19	Surgical Intervention	A070504	Premature Discharge of Battery	17/12/2020	17/12/2020
							F1903	Device Explantation	A070504	Premature Discharge of Battery	17/12/2020	17/12/2020
					CT233	Surgical	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	17/12/2020	17/12/2020
							F12	Serious Injury/ Illness/ Impairment	A070504	Premature Discharge of Battery	17/12/2020	17/12/2020
							F19	Surgical Intervention	A070504	Premature Discharge of Battery	17/12/2020	17/12/2020
							F1903	Device Explantation	A070504	Premature Discharge of Battery	17/12/2020	17/12/2020
					CT2406	Implantable	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	17/12/2020	17/12/2020
							F12	Serious Injury/ Illness/ Impairment	A070504	Premature Discharge of Battery	17/12/2020	17/12/2020
							F19	Surgical Intervention	A070504	Premature Discharge of Battery	17/12/2020	17/12/2020
							F1903	Device Explantation	A070504	Premature Discharge of Battery	17/12/2020	17/12/2020
					CT2407	Active implantable	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	17/12/2020	17/12/2020
							F12	Serious Injury/ Illness/ Impairment	A070504	Premature Discharge of Battery	17/12/2020	17/12/2020
							F19	Surgical Intervention	A070504	Premature Discharge of Battery	17/12/2020	17/12/2020
							F1903	Device Explantation	A070504	Premature Discharge of Battery	17/12/2020	17/12/2020
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	17/12/2020	17/12/2020
							F12	Serious Injury/ Illness/ Impairment	A070504	Premature Discharge of Battery	17/12/2020	17/12/2020
							F19	Surgical Intervention	A070504	Premature Discharge of Battery	17/12/2020	17/12/2020
							F1903	Device Explantation	A070504	Premature Discharge of Battery	17/12/2020	17/12/2020
					CT2727	Active implant devices/ancillaries	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	17/12/2020	17/12/2020
							F12	Serious Injury/ Illness/ Impairment	A070504	Premature Discharge of Battery	17/12/2020	17/12/2020
							F19	Surgical Intervention	A070504	Premature Discharge of Battery	17/12/2020	17/12/2020
							F1903	Device Explantation	A070504	Premature Discharge of Battery	17/12/2020	17/12/2020
					CT2750	Implantable cardiac electrical stimulators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	17/12/2020	17/12/2020
							F12	Serious Injury/ Illness/ Impairment	A070504	Premature Discharge of Battery	17/12/2020	17/12/2020
							F19	Surgical Intervention	A070504	Premature Discharge of Battery	17/12/2020	17/12/2020
							F1903	Device Explantation	A070504	Premature Discharge of Battery	17/12/2020	17/12/2020
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	17/12/2020	17/12/2020
							F12	Serious Injury/ Illness/ Impairment	A070504	Premature Discharge of Battery	17/12/2020	17/12/2020
							F19	Surgical Intervention	A070504	Premature Discharge of Battery	17/12/2020	17/12/2020
							F1903	Device Explantation	A070504	Premature Discharge of Battery	17/12/2020	17/12/2020
					CT2985	Active	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	17/12/2020	17/12/2020
							F12	Serious Injury/ Illness/ Impairment	A070504	Premature Discharge of Battery	17/12/2020	17/12/2020
							F19	Surgical Intervention	A070504	Premature Discharge of Battery	17/12/2020	17/12/2020
							F1903	Device Explantation	A070504	Premature Discharge of Battery	17/12/2020	17/12/2020
					CT3001	Implantable metal	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	17/12/2020	17/12/2020
							F12	Serious Injury/ Illness/ Impairment	A070504	Premature Discharge of Battery	17/12/2020	17/12/2020
							F19	Surgical Intervention	A070504	Premature Discharge of Battery	17/12/2020	17/12/2020
							F1903	Device Explantation	A070504	Premature Discharge of Battery	17/12/2020	17/12/2020
					CT3002	Implantable synthetic polymer	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	17/12/2020	17/12/2020
							F12	Serious Injury/ Illness/ Impairment	A070504	Premature Discharge of Battery	17/12/2020	17/12/2020
							F19	Surgical Intervention	A070504	Premature Discharge of Battery	17/12/2020	17/12/2020
							F1903	Device Explantation	A070504	Premature Discharge of Battery	17/12/2020	17/12/2020
					CT3006	Implantable inorganic material	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	17/12/2020	17/12/2020
							F12	Serious Injury/ Illness/ Impairment	A070504	Premature Discharge of Battery	17/12/2020	17/12/2020
							F19	Surgical Intervention	A070504	Premature Discharge of Battery	17/12/2020	17/12/2020
							F1903	Device Explantation	A070504	Premature Discharge of Battery	17/12/2020	17/12/2020
					CT314	Electrically powered	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	17/12/2020	17/12/2020
							F12	Serious Injury/ Illness/ Impairment	A070504	Premature Discharge of Battery	17/12/2020	17/12/2020
							F19	Surgical Intervention	A070504	Premature Discharge of Battery	17/12/2020	17/12/2020
							F1903	Device Explantation	A070504	Premature Discharge of Battery	17/12/2020	17/12/2020
					CT321	Long-term surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	17/12/2020	17/12/2020
							F12	Serious Injury/ Illness/ Impairment	A070504	Premature Discharge of Battery	17/12/2020	17/12/2020
							F19	Surgical Intervention	A070504	Premature Discharge of Battery	17/12/2020	17/12/2020
							F1903	Device Explantation	A070504	Premature Discharge of Battery	17/12/2020	17/12/2020
					CT449	Generators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	17/12/2020	17/12/2020
							F12	Serious Injury/ Illness/ Impairment	A070504	Premature Discharge of Battery	17/12/2020	17/12/2020
							F19	Surgical Intervention	A070504	Premature Discharge of Battery	17/12/2020	17/12/2020
							F1903	Device Explantation	A070504	Premature Discharge of Battery	17/12/2020	17/12/2020
					CT981	Single use	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	17/12/2020	17/12/2020
							F12	Serious Injury/ Illness/ Impairment	A070504	Premature Discharge of Battery	17/12/2020	17/12/2020
							F19	Surgical Intervention	A070504	Premature Discharge of Battery	17/12/2020	17/12/2020
							F1903	Device Explantation	A070504	Premature Discharge of Battery	17/12/2020	17/12/2020
					CT983	Surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	17/12/2020	17/12/2020
							F12	Serious Injury/ Illness/ Impairment	A070504	Premature Discharge of Battery	17/12/2020	17/12/2020
							F19	Surgical Intervention	A070504	Premature Discharge of Battery	17/12/2020	17/12/2020
							F1903	Device Explantation	A070504	Premature Discharge of Battery	17/12/2020	17/12/2020
10846	2021	All other reportable incidents	47265	Dual-chamber implantable pacemaker, rate-responsive	CT141	Electrodes/Leads and associated devices	E2103	Device Overstimulation of Tissue	A0705	Battery Problem	21/05/2021	21/05/2021
									A070504	Premature Discharge of Battery	21/05/2021	21/05/2021
									A072201	High Impedance	21/05/2021	21/05/2021
							E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem	21/05/2021	21/05/2021
									A070504	Premature Discharge of Battery	21/05/2021	21/05/2021
									A072201	High Impedance	21/05/2021	21/05/2021
							F08	Hospitalization or Prolonged Hospitalization	A0705	Battery Problem	21/05/2021	21/05/2021
									A070504	Premature Discharge of Battery	21/05/2021	21/05/2021
									A072201	High Impedance	21/05/2021	21/05/2021
							F12	Serious Injury/ Illness/ Impairment	A0705	Battery Problem	21/05/2021	21/05/2021
									A070504	Premature Discharge of Battery	21/05/2021	21/05/2021
									A072201	High Impedance	21/05/2021	21/05/2021

							F1905	Device Revision or Replacement	A0705	Battery Problem	21/05/2021	21/05/2021
									A070504	Premature Discharge of Battery	21/05/2021	21/05/2021
									A072201	High Impedance	21/05/2021	21/05/2021
						CT145	Cardiovascular devices	E2103	Device Overstimulation of Tissue	A0705	Battery Problem	21/05/2021	21/05/2021
										A070504	Premature Discharge of Battery	21/05/2021	21/05/2021
										A072201	High Impedance	21/05/2021	21/05/2021
								E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem	21/05/2021	21/05/2021
										A070504	Premature Discharge of Battery	21/05/2021	21/05/2021
										A072201	High Impedance	21/05/2021	21/05/2021
								F08	Hospitalization or Prolonged Hospitalization	A0705	Battery Problem	21/05/2021	21/05/2021
										A070504	Premature Discharge of Battery	21/05/2021	21/05/2021
										A072201	High Impedance	21/05/2021	21/05/2021
								F12	Serious Injury/ Illness/ Impairment	A0705	Battery Problem	21/05/2021	21/05/2021
										A070504	Premature Discharge of Battery	21/05/2021	21/05/2021
										A072201	High Impedance	21/05/2021	21/05/2021
								F1905	Device Revision or Replacement	A0705	Battery Problem	21/05/2021	21/05/2021
										A070504	Premature Discharge of Battery	21/05/2021	21/05/2021
										A072201	High Impedance	21/05/2021	21/05/2021
						CT1684	Cardiac pacemakers and associated devices	E2103	Device Overstimulation of Tissue	A0705	Battery Problem	21/05/2021	21/05/2021
										A070504	Premature Discharge of Battery	21/05/2021	21/05/2021
										A072201	High Impedance	21/05/2021	21/05/2021
								E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem	21/05/2021	21/05/2021
										A070504	Premature Discharge of Battery	21/05/2021	21/05/2021
										A072201	High Impedance	21/05/2021	21/05/2021
								F08	Hospitalization or Prolonged Hospitalization	A0705	Battery Problem	21/05/2021	21/05/2021
										A070504	Premature Discharge of Battery	21/05/2021	21/05/2021
										A072201	High Impedance	21/05/2021	21/05/2021
								F12	Serious Injury/ Illness/ Impairment	A0705	Battery Problem	21/05/2021	21/05/2021
										A070504	Premature Discharge of Battery	21/05/2021	21/05/2021
										A072201	High Impedance	21/05/2021	21/05/2021
								F1905	Device Revision or Replacement	A0705	Battery Problem	21/05/2021	21/05/2021
										A070504	Premature Discharge of Battery	21/05/2021	21/05/2021
										A072201	High Impedance	21/05/2021	21/05/2021
						CT1685	Cardiac pacemakers	E2103	Device Overstimulation of Tissue	A0705	Battery Problem	21/05/2021	21/05/2021
										A070504	Premature Discharge of Battery	21/05/2021	21/05/2021
										A072201	High Impedance	21/05/2021	21/05/2021
								E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem	21/05/2021	21/05/2021
										A070504	Premature Discharge of Battery	21/05/2021	21/05/2021
										A072201	High Impedance	21/05/2021	21/05/2021
								F08	Hospitalization or Prolonged Hospitalization	A0705	Battery Problem	21/05/2021	21/05/2021
										A070504	Premature Discharge of Battery	21/05/2021	21/05/2021
										A072201	High Impedance	21/05/2021	21/05/2021
								F12	Serious Injury/ Illness/ Impairment	A0705	Battery Problem	21/05/2021	21/05/2021
										A070504	Premature Discharge of Battery	21/05/2021	21/05/2021
										A072201	High Impedance	21/05/2021	21/05/2021
								F1905	Device Revision or Replacement	A0705	Battery Problem	21/05/2021	21/05/2021
										A070504	Premature Discharge of Battery	21/05/2021	21/05/2021
										A072201	High Impedance	21/05/2021	21/05/2021
						CT233	Surgical	E2103	Device Overstimulation of Tissue	A0705	Battery Problem	21/05/2021	21/05/2021
										A070504	Premature Discharge of Battery	21/05/2021	21/05/2021
										A072201	High Impedance	21/05/2021	21/05/2021
								E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem	21/05/2021	21/05/2021
										A070504	Premature Discharge of Battery	21/05/2021	21/05/2021
										A072201	High Impedance	21/05/2021	21/05/2021
								F08	Hospitalization or Prolonged Hospitalization	A0705	Battery Problem	21/05/2021	21/05/2021
										A070504	Premature Discharge of Battery	21/05/2021	21/05/2021
										A072201	High Impedance	21/05/2021	21/05/2021
								F12	Serious Injury/ Illness/ Impairment	A0705	Battery Problem	21/05/2021	21/05/2021
										A070504	Premature Discharge of Battery	21/05/2021	21/05/2021
										A072201	High Impedance	21/05/2021	21/05/2021
								F1905	Device Revision or Replacement	A0705	Battery Problem	21/05/2021	21/05/2021
										A070504	Premature Discharge of Battery	21/05/2021	21/05/2021
										A072201	High Impedance	21/05/2021	21/05/2021
						CT2406	Implantable	E2103	Device Overstimulation of Tissue	A0705	Battery Problem	21/05/2021	21/05/2021
										A070504	Premature Discharge of Battery	21/05/2021	21/05/2021
										A072201	High Impedance	21/05/2021	21/05/2021
								E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem	21/05/2021	21/05/2021
										A070504	Premature Discharge of Battery	21/05/2021	21/05/2021
										A072201	High Impedance	21/05/2021	21/05/2021
								F08	Hospitalization or Prolonged Hospitalization	A0705	Battery Problem	21/05/2021	21/05/2021
										A070504	Premature Discharge of Battery	21/05/2021	21/05/2021
										A072201	High Impedance	21/05/2021	21/05/2021
								F12	Serious Injury/ Illness/ Impairment	A0705	Battery Problem	21/05/2021	21/05/2021
										A070504	Premature Discharge of Battery	21/05/2021	21/05/2021
										A072201	High Impedance	21/05/2021	21/05/2021
								F1905	Device Revision or Replacement	A0705	Battery Problem	21/05/2021	21/05/2021
										A070504	Premature Discharge of Battery	21/05/2021	21/05/2021
										A072201	High Impedance	21/05/2021	21/05/2021
						CT2407	Active implantable	E2103	Device Overstimulation of Tissue	A0705	Battery Problem	21/05/2021	21/05/2021
										A070504	Premature Discharge of Battery	21/05/2021	21/05/2021
										A072201	High Impedance	21/05/2021	21/05/2021
								E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem	21/05/2021	21/05/2021
										A070504	Premature Discharge of Battery	21/05/2021	21/05/2021
										A072201	High Impedance	21/05/2021	21/05/2021
								F08	Hospitalization or Prolonged Hospitalization	A0705	Battery Problem	21/05/2021	21/05/2021
										A070504	Premature Discharge of Battery	21/05/2021	21/05/2021
										A072201	High Impedance	21/05/2021	21/05/2021
								F12	Serious Injury/ Illness/ Impairment	A0705	Battery Problem	21/05/2021	21/05/2021

								A070504	Premature Discharge of Battery	21/05/2021	21/05/2021
								A072201	High Impedance	21/05/2021	21/05/2021
							F1905	Device Revision or Replacement	A0705	Battery Problem	21/05/2021	21/05/2021
									A070504	Premature Discharge of Battery	21/05/2021	21/05/2021
									A072201	High Impedance	21/05/2021	21/05/2021
					CT257	Implantable cardiac pacemakers	E2103	Device Overstimulation of Tissue	A0705	Battery Problem	21/05/2021	21/05/2021
									A070504	Premature Discharge of Battery	21/05/2021	21/05/2021
									A072201	High Impedance	21/05/2021	21/05/2021
							E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem	21/05/2021	21/05/2021
									A070504	Premature Discharge of Battery	21/05/2021	21/05/2021
									A072201	High Impedance	21/05/2021	21/05/2021
							F08	Hospitalization or Prolonged Hospitalization	A0705	Battery Problem	21/05/2021	21/05/2021
									A070504	Premature Discharge of Battery	21/05/2021	21/05/2021
									A072201	High Impedance	21/05/2021	21/05/2021
							F12	Serious Injury/ Illness/ Impairment	A0705	Battery Problem	21/05/2021	21/05/2021
									A070504	Premature Discharge of Battery	21/05/2021	21/05/2021
									A072201	High Impedance	21/05/2021	21/05/2021
							F1905	Device Revision or Replacement	A0705	Battery Problem	21/05/2021	21/05/2021
									A070504	Premature Discharge of Battery	21/05/2021	21/05/2021
									A072201	High Impedance	21/05/2021	21/05/2021
					CT2727	Active implant devices/ancillaries	E2103	Device Overstimulation of Tissue	A0705	Battery Problem	21/05/2021	21/05/2021
									A070504	Premature Discharge of Battery	21/05/2021	21/05/2021
									A072201	High Impedance	21/05/2021	21/05/2021
							E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem	21/05/2021	21/05/2021
									A070504	Premature Discharge of Battery	21/05/2021	21/05/2021
									A072201	High Impedance	21/05/2021	21/05/2021
							F08	Hospitalization or Prolonged Hospitalization	A0705	Battery Problem	21/05/2021	21/05/2021
									A070504	Premature Discharge of Battery	21/05/2021	21/05/2021
									A072201	High Impedance	21/05/2021	21/05/2021
							F12	Serious Injury/ Illness/ Impairment	A0705	Battery Problem	21/05/2021	21/05/2021
									A070504	Premature Discharge of Battery	21/05/2021	21/05/2021
									A072201	High Impedance	21/05/2021	21/05/2021
							F1905	Device Revision or Replacement	A0705	Battery Problem	21/05/2021	21/05/2021
									A070504	Premature Discharge of Battery	21/05/2021	21/05/2021
									A072201	High Impedance	21/05/2021	21/05/2021
					CT2750	Implantable cardiac electrical stimulators and associated devices	E2103	Device Overstimulation of Tissue	A0705	Battery Problem	21/05/2021	21/05/2021
									A070504	Premature Discharge of Battery	21/05/2021	21/05/2021
									A072201	High Impedance	21/05/2021	21/05/2021
							E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem	21/05/2021	21/05/2021
									A070504	Premature Discharge of Battery	21/05/2021	21/05/2021
									A072201	High Impedance	21/05/2021	21/05/2021
							F08	Hospitalization or Prolonged Hospitalization	A0705	Battery Problem	21/05/2021	21/05/2021
									A070504	Premature Discharge of Battery	21/05/2021	21/05/2021
									A072201	High Impedance	21/05/2021	21/05/2021
							F12	Serious Injury/ Illness/ Impairment	A0705	Battery Problem	21/05/2021	21/05/2021
									A070504	Premature Discharge of Battery	21/05/2021	21/05/2021
									A072201	High Impedance	21/05/2021	21/05/2021
							F1905	Device Revision or Replacement	A0705	Battery Problem	21/05/2021	21/05/2021
									A070504	Premature Discharge of Battery	21/05/2021	21/05/2021
									A072201	High Impedance	21/05/2021	21/05/2021
					CT2866	Implantable cardiac electrical stimulators	E2103	Device Overstimulation of Tissue	A0705	Battery Problem	21/05/2021	21/05/2021
									A070504	Premature Discharge of Battery	21/05/2021	21/05/2021
									A072201	High Impedance	21/05/2021	21/05/2021
							E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem	21/05/2021	21/05/2021
									A070504	Premature Discharge of Battery	21/05/2021	21/05/2021
									A072201	High Impedance	21/05/2021	21/05/2021
							F08	Hospitalization or Prolonged Hospitalization	A0705	Battery Problem	21/05/2021	21/05/2021
									A070504	Premature Discharge of Battery	21/05/2021	21/05/2021
									A072201	High Impedance	21/05/2021	21/05/2021
							F12	Serious Injury/ Illness/ Impairment	A0705	Battery Problem	21/05/2021	21/05/2021
									A070504	Premature Discharge of Battery	21/05/2021	21/05/2021
									A072201	High Impedance	21/05/2021	21/05/2021
							F1905	Device Revision or Replacement	A0705	Battery Problem	21/05/2021	21/05/2021
									A070504	Premature Discharge of Battery	21/05/2021	21/05/2021
									A072201	High Impedance	21/05/2021	21/05/2021
					CT2985	Active	E2103	Device Overstimulation of Tissue	A0705	Battery Problem	21/05/2021	21/05/2021
									A070504	Premature Discharge of Battery	21/05/2021	21/05/2021
									A072201	High Impedance	21/05/2021	21/05/2021
							E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem	21/05/2021	21/05/2021
									A070504	Premature Discharge of Battery	21/05/2021	21/05/2021
									A072201	High Impedance	21/05/2021	21/05/2021
							F08	Hospitalization or Prolonged Hospitalization	A0705	Battery Problem	21/05/2021	21/05/2021
									A070504	Premature Discharge of Battery	21/05/2021	21/05/2021
									A072201	High Impedance	21/05/2021	21/05/2021
							F12	Serious Injury/ Illness/ Impairment	A0705	Battery Problem	21/05/2021	21/05/2021
									A070504	Premature Discharge of Battery	21/05/2021	21/05/2021
									A072201	High Impedance	21/05/2021	21/05/2021
							F1905	Device Revision or Replacement	A0705	Battery Problem	21/05/2021	21/05/2021
									A070504	Premature Discharge of Battery	21/05/2021	21/05/2021
									A072201	High Impedance	21/05/2021	21/05/2021
					CT3001	Implantable metal	E2103	Device Overstimulation of Tissue	A0705	Battery Problem	21/05/2021	21/05/2021
									A070504	Premature Discharge of Battery	21/05/2021	21/05/2021
									A072201	High Impedance	21/05/2021	21/05/2021
							E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem	21/05/2021	21/05/2021
									A070504	Premature Discharge of Battery	21/05/2021	21/05/2021
									A072201	High Impedance	21/05/2021	21/05/2021
							F08	Hospitalization or Prolonged Hospitalization	A0705	Battery Problem	21/05/2021	21/05/2021
									A070504	Premature Discharge of Battery	21/05/2021	21/05/2021

								A072201	High Impedance	21/05/2021	21/05/2021
						F12	Serious Injury/ Illness/ Impairment	A0705	Battery Problem	21/05/2021	21/05/2021
								A070504	Premature Discharge of Battery	21/05/2021	21/05/2021
								A072201	High Impedance	21/05/2021	21/05/2021
						F1905	Device Revision or Replacement	A0705	Battery Problem	21/05/2021	21/05/2021
								A070504	Premature Discharge of Battery	21/05/2021	21/05/2021
								A072201	High Impedance	21/05/2021	21/05/2021
					CT3002	Implantable synthetic polymer	E2103	Device Overstimulation of Tissue	A0705	Battery Problem	21/05/2021	21/05/2021
									A070504	Premature Discharge of Battery	21/05/2021	21/05/2021
									A072201	High Impedance	21/05/2021	21/05/2021
							E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem	21/05/2021	21/05/2021
									A070504	Premature Discharge of Battery	21/05/2021	21/05/2021
									A072201	High Impedance	21/05/2021	21/05/2021
							F08	Hospitalization or Prolonged Hospitalization	A0705	Battery Problem	21/05/2021	21/05/2021
									A070504	Premature Discharge of Battery	21/05/2021	21/05/2021
									A072201	High Impedance	21/05/2021	21/05/2021
							F12	Serious Injury/ Illness/ Impairment	A0705	Battery Problem	21/05/2021	21/05/2021
									A070504	Premature Discharge of Battery	21/05/2021	21/05/2021
									A072201	High Impedance	21/05/2021	21/05/2021
							F1905	Device Revision or Replacement	A0705	Battery Problem	21/05/2021	21/05/2021
									A070504	Premature Discharge of Battery	21/05/2021	21/05/2021
									A072201	High Impedance	21/05/2021	21/05/2021
					CT3006	Implantable inorganic material	E2103	Device Overstimulation of Tissue	A0705	Battery Problem	21/05/2021	21/05/2021
									A070504	Premature Discharge of Battery	21/05/2021	21/05/2021
									A072201	High Impedance	21/05/2021	21/05/2021
							E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem	21/05/2021	21/05/2021
									A070504	Premature Discharge of Battery	21/05/2021	21/05/2021
									A072201	High Impedance	21/05/2021	21/05/2021
							F08	Hospitalization or Prolonged Hospitalization	A0705	Battery Problem	21/05/2021	21/05/2021
									A070504	Premature Discharge of Battery	21/05/2021	21/05/2021
									A072201	High Impedance	21/05/2021	21/05/2021
							F12	Serious Injury/ Illness/ Impairment	A0705	Battery Problem	21/05/2021	21/05/2021
									A070504	Premature Discharge of Battery	21/05/2021	21/05/2021
									A072201	High Impedance	21/05/2021	21/05/2021
							F1905	Device Revision or Replacement	A0705	Battery Problem	21/05/2021	21/05/2021
									A070504	Premature Discharge of Battery	21/05/2021	21/05/2021
									A072201	High Impedance	21/05/2021	21/05/2021
					CT314	Electrically powered	E2103	Device Overstimulation of Tissue	A0705	Battery Problem	21/05/2021	21/05/2021
									A070504	Premature Discharge of Battery	21/05/2021	21/05/2021
									A072201	High Impedance	21/05/2021	21/05/2021
							E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem	21/05/2021	21/05/2021
									A070504	Premature Discharge of Battery	21/05/2021	21/05/2021
									A072201	High Impedance	21/05/2021	21/05/2021
							F08	Hospitalization or Prolonged Hospitalization	A0705	Battery Problem	21/05/2021	21/05/2021
									A070504	Premature Discharge of Battery	21/05/2021	21/05/2021
									A072201	High Impedance	21/05/2021	21/05/2021
							F12	Serious Injury/ Illness/ Impairment	A0705	Battery Problem	21/05/2021	21/05/2021
									A070504	Premature Discharge of Battery	21/05/2021	21/05/2021
									A072201	High Impedance	21/05/2021	21/05/2021
							F1905	Device Revision or Replacement	A0705	Battery Problem	21/05/2021	21/05/2021
									A070504	Premature Discharge of Battery	21/05/2021	21/05/2021
									A072201	High Impedance	21/05/2021	21/05/2021
					CT321	Long-term surgical invasive	E2103	Device Overstimulation of Tissue	A0705	Battery Problem	21/05/2021	21/05/2021
									A070504	Premature Discharge of Battery	21/05/2021	21/05/2021
									A072201	High Impedance	21/05/2021	21/05/2021
							E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem	21/05/2021	21/05/2021
									A070504	Premature Discharge of Battery	21/05/2021	21/05/2021
									A072201	High Impedance	21/05/2021	21/05/2021
							F08	Hospitalization or Prolonged Hospitalization	A0705	Battery Problem	21/05/2021	21/05/2021
									A070504	Premature Discharge of Battery	21/05/2021	21/05/2021
									A072201	High Impedance	21/05/2021	21/05/2021
							F12	Serious Injury/ Illness/ Impairment	A0705	Battery Problem	21/05/2021	21/05/2021
									A070504	Premature Discharge of Battery	21/05/2021	21/05/2021
									A072201	High Impedance	21/05/2021	21/05/2021
							F1905	Device Revision or Replacement	A0705	Battery Problem	21/05/2021	21/05/2021
									A070504	Premature Discharge of Battery	21/05/2021	21/05/2021
									A072201	High Impedance	21/05/2021	21/05/2021
					CT449	Generators and associated devices	E2103	Device Overstimulation of Tissue	A0705	Battery Problem	21/05/2021	21/05/2021
									A070504	Premature Discharge of Battery	21/05/2021	21/05/2021
									A072201	High Impedance	21/05/2021	21/05/2021
							E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem	21/05/2021	21/05/2021
									A070504	Premature Discharge of Battery	21/05/2021	21/05/2021
									A072201	High Impedance	21/05/2021	21/05/2021
							F08	Hospitalization or Prolonged Hospitalization	A0705	Battery Problem	21/05/2021	21/05/2021
									A070504	Premature Discharge of Battery	21/05/2021	21/05/2021
									A072201	High Impedance	21/05/2021	21/05/2021
							F12	Serious Injury/ Illness/ Impairment	A0705	Battery Problem	21/05/2021	21/05/2021
									A070504	Premature Discharge of Battery	21/05/2021	21/05/2021
									A072201	High Impedance	21/05/2021	21/05/2021
							F1905	Device Revision or Replacement	A0705	Battery Problem	21/05/2021	21/05/2021
									A070504	Premature Discharge of Battery	21/05/2021	21/05/2021
									A072201	High Impedance	21/05/2021	21/05/2021
					CT981	Single use	E2103	Device Overstimulation of Tissue	A0705	Battery Problem	21/05/2021	21/05/2021
									A070504	Premature Discharge of Battery	21/05/2021	21/05/2021
									A072201	High Impedance	21/05/2021	21/05/2021
							E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem	21/05/2021	21/05/2021
									A070504	Premature Discharge of Battery	21/05/2021	21/05/2021
									A072201	High Impedance	21/05/2021	21/05/2021

						F08	Hospitalization or Prolonged Hospitalization	A0705	Battery Problem	21/05/2021	21/05/2021
								A070504	Premature Discharge of Battery	21/05/2021	21/05/2021
								A072201	High Impedance	21/05/2021	21/05/2021
						F12	Serious Injury/ Illness/ Impairment	A0705	Battery Problem	21/05/2021	21/05/2021
								A070504	Premature Discharge of Battery	21/05/2021	21/05/2021
								A072201	High Impedance	21/05/2021	21/05/2021
						F1905	Device Revision or Replacement	A0705	Battery Problem	21/05/2021	21/05/2021
								A070504	Premature Discharge of Battery	21/05/2021	21/05/2021
								A072201	High Impedance	21/05/2021	21/05/2021
				CT983	Surgical invasive	E2103	Device Overstimulation of Tissue	A0705	Battery Problem	21/05/2021	21/05/2021
								A070504	Premature Discharge of Battery	21/05/2021	21/05/2021
								A072201	High Impedance	21/05/2021	21/05/2021
						E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem	21/05/2021	21/05/2021
								A070504	Premature Discharge of Battery	21/05/2021	21/05/2021
								A072201	High Impedance	21/05/2021	21/05/2021
						F08	Hospitalization or Prolonged Hospitalization	A0705	Battery Problem	21/05/2021	21/05/2021
								A070504	Premature Discharge of Battery	21/05/2021	21/05/2021
								A072201	High Impedance	21/05/2021	21/05/2021
						F12	Serious Injury/ Illness/ Impairment	A0705	Battery Problem	21/05/2021	21/05/2021
								A070504	Premature Discharge of Battery	21/05/2021	21/05/2021
								A072201	High Impedance	21/05/2021	21/05/2021
						F1905	Device Revision or Replacement	A0705	Battery Problem	21/05/2021	21/05/2021
								A070504	Premature Discharge of Battery	21/05/2021	21/05/2021
								A072201	High Impedance	21/05/2021	21/05/2021
10847	2021	All other reportable incidents	47265	Dual-chamber implantable pacemaker, rate-responsive	CT141	Electrodes/Leads and associated devices	E060109	Tachycardia	A070910	Undersensing		
									A0712	Pacing Problem		
									A071202	Inaccurate Synchronization		
							F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A070910	Undersensing		
									A0712	Pacing Problem		
									A071202	Inaccurate Synchronization		
							F11	Non-Serious Injury/ Illness/ Impairment	A070910	Undersensing		
									A0712	Pacing Problem		
									A071202	Inaccurate Synchronization		
					CT145	Cardiovascular devices	E060109	Tachycardia	A070910	Undersensing		
									A0712	Pacing Problem		
									A071202	Inaccurate Synchronization		
							F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A070910	Undersensing		
									A0712	Pacing Problem		
									A071202	Inaccurate Synchronization		
							F11	Non-Serious Injury/ Illness/ Impairment	A070910	Undersensing		
									A0712	Pacing Problem		
									A071202	Inaccurate Synchronization		
					CT1684	Cardiac pacemakers and associated devices	E060109	Tachycardia	A070910	Undersensing		
									A0712	Pacing Problem		
									A071202	Inaccurate Synchronization		
							F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A070910	Undersensing		
									A0712	Pacing Problem		
									A071202	Inaccurate Synchronization		
							F11	Non-Serious Injury/ Illness/ Impairment	A070910	Undersensing		
									A0712	Pacing Problem		
									A071202	Inaccurate Synchronization		
					CT1685	Cardiac pacemakers	E060109	Tachycardia	A070910	Undersensing		
									A0712	Pacing Problem		
									A071202	Inaccurate Synchronization		
							F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A070910	Undersensing		
									A0712	Pacing Problem		
									A071202	Inaccurate Synchronization		
							F11	Non-Serious Injury/ Illness/ Impairment	A070910	Undersensing		
									A0712	Pacing Problem		
									A071202	Inaccurate Synchronization		
					CT233	Surgical	E060109	Tachycardia	A070910	Undersensing		
									A0712	Pacing Problem		
									A071202	Inaccurate Synchronization		
							F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A070910	Undersensing		
									A0712	Pacing Problem		
									A071202	Inaccurate Synchronization		
							F11	Non-Serious Injury/ Illness/ Impairment	A070910	Undersensing		
									A0712	Pacing Problem		
									A071202	Inaccurate Synchronization		
					CT2406	Implantable	E060109	Tachycardia	A070910	Undersensing		
									A0712	Pacing Problem		
									A071202	Inaccurate Synchronization		
							F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A070910	Undersensing		
									A0712	Pacing Problem		
									A071202	Inaccurate Synchronization		
							F11	Non-Serious Injury/ Illness/ Impairment	A070910	Undersensing		
									A0712	Pacing Problem		
									A071202	Inaccurate Synchronization		
					CT2407	Active implantable	E060109	Tachycardia	A070910	Undersensing		
									A0712	Pacing Problem		
									A071202	Inaccurate Synchronization		
							F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A070910	Undersensing		
									A0712	Pacing Problem		
									A071202	Inaccurate Synchronization		
							F11	Non-Serious Injury/ Illness/ Impairment	A070910	Undersensing		
									A0712	Pacing Problem		
									A071202	Inaccurate Synchronization		
					CT257	Implantable cardiac pacemakers	E060109	Tachycardia	A070910	Undersensing		

									A0712	Pacing Problem		
									A071202	Inaccurate Synchronization		
							F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A070910	Undersensing		
									A0712	Pacing Problem		
									A071202	Inaccurate Synchronization		
							F11	Non-Serious Injury/ Illness/ Impairment	A070910	Undersensing		
									A0712	Pacing Problem		
									A071202	Inaccurate Synchronization		
					CT2727	Active implant devices/ancillaries	E060109	Tachycardia	A070910	Undersensing		
									A0712	Pacing Problem		
									A071202	Inaccurate Synchronization		
							F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A070910	Undersensing		
									A0712	Pacing Problem		
									A071202	Inaccurate Synchronization		
							F11	Non-Serious Injury/ Illness/ Impairment	A070910	Undersensing		
									A0712	Pacing Problem		
									A071202	Inaccurate Synchronization		
					CT2750	Implantable cardiac electrical stimulators and associated devices	E060109	Tachycardia	A070910	Undersensing		
									A0712	Pacing Problem		
									A071202	Inaccurate Synchronization		
							F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A070910	Undersensing		
									A0712	Pacing Problem		
									A071202	Inaccurate Synchronization		
							F11	Non-Serious Injury/ Illness/ Impairment	A070910	Undersensing		
									A0712	Pacing Problem		
									A071202	Inaccurate Synchronization		
					CT2866	Implantable cardiac electrical stimulators	E060109	Tachycardia	A070910	Undersensing		
									A0712	Pacing Problem		
									A071202	Inaccurate Synchronization		
							F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A070910	Undersensing		
									A0712	Pacing Problem		
									A071202	Inaccurate Synchronization		
							F11	Non-Serious Injury/ Illness/ Impairment	A070910	Undersensing		
									A0712	Pacing Problem		
									A071202	Inaccurate Synchronization		
					CT2985	Active	E060109	Tachycardia	A070910	Undersensing		
									A0712	Pacing Problem		
									A071202	Inaccurate Synchronization		
							F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A070910	Undersensing		
									A0712	Pacing Problem		
									A071202	Inaccurate Synchronization		
							F11	Non-Serious Injury/ Illness/ Impairment	A070910	Undersensing		
									A0712	Pacing Problem		
									A071202	Inaccurate Synchronization		
					CT3001	Implantable metal	E060109	Tachycardia	A070910	Undersensing		
									A0712	Pacing Problem		
									A071202	Inaccurate Synchronization		
							F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A070910	Undersensing		
									A0712	Pacing Problem		
									A071202	Inaccurate Synchronization		
							F11	Non-Serious Injury/ Illness/ Impairment	A070910	Undersensing		
									A0712	Pacing Problem		
									A071202	Inaccurate Synchronization		
					CT3002	Implantable synthetic polymer	E060109	Tachycardia	A070910	Undersensing		
									A0712	Pacing Problem		
									A071202	Inaccurate Synchronization		
							F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A070910	Undersensing		
									A0712	Pacing Problem		
									A071202	Inaccurate Synchronization		
							F11	Non-Serious Injury/ Illness/ Impairment	A070910	Undersensing		
									A0712	Pacing Problem		
									A071202	Inaccurate Synchronization		
					CT3006	Implantable inorganic material	E060109	Tachycardia	A070910	Undersensing		
									A0712	Pacing Problem		
									A071202	Inaccurate Synchronization		
							F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A070910	Undersensing		
									A0712	Pacing Problem		
									A071202	Inaccurate Synchronization		
							F11	Non-Serious Injury/ Illness/ Impairment	A070910	Undersensing		
									A0712	Pacing Problem		
									A071202	Inaccurate Synchronization		
					CT314	Electrically powered	E060109	Tachycardia	A070910	Undersensing		
									A0712	Pacing Problem		
									A071202	Inaccurate Synchronization		
							F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A070910	Undersensing		
									A0712	Pacing Problem		
									A071202	Inaccurate Synchronization		
							F11	Non-Serious Injury/ Illness/ Impairment	A070910	Undersensing		
									A0712	Pacing Problem		
									A071202	Inaccurate Synchronization		
					CT321	Long-term surgical invasive	E060109	Tachycardia	A070910	Undersensing		
									A0712	Pacing Problem		
									A071202	Inaccurate Synchronization		
							F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A070910	Undersensing		
									A0712	Pacing Problem		
									A071202	Inaccurate Synchronization		
							F11	Non-Serious Injury/ Illness/ Impairment	A070910	Undersensing		
									A0712	Pacing Problem		

									A071202	Inaccurate Synchronization		
					CT449	Generators and associated devices	E060109	Tachycardia	A070910	Undersensing		
									A0712	Pacing Problem		
									A071202	Inaccurate Synchronization		
							F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A070910	Undersensing		
									A0712	Pacing Problem		
									A071202	Inaccurate Synchronization		
							F11	Non-Serious Injury/ Illness/ Impairment	A070910	Undersensing		
									A0712	Pacing Problem		
									A071202	Inaccurate Synchronization		
					CT981	Single use	E060109	Tachycardia	A070910	Undersensing		
									A0712	Pacing Problem		
									A071202	Inaccurate Synchronization		
							F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A070910	Undersensing		
									A0712	Pacing Problem		
									A071202	Inaccurate Synchronization		
							F11	Non-Serious Injury/ Illness/ Impairment	A070910	Undersensing		
									A0712	Pacing Problem		
									A071202	Inaccurate Synchronization		
					CT983	Surgical invasive	E060109	Tachycardia	A070910	Undersensing		
									A0712	Pacing Problem		
									A071202	Inaccurate Synchronization		
							F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A070910	Undersensing		
									A0712	Pacing Problem		
									A071202	Inaccurate Synchronization		
							F11	Non-Serious Injury/ Illness/ Impairment	A070910	Undersensing		
									A0712	Pacing Problem		
									A071202	Inaccurate Synchronization		
10848	2021	All other reportable incidents	35852	Single-chamber implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem	25/03/2021	25/03/2021
									A070504	Premature Discharge of Battery	25/03/2021	25/03/2021
							F19	Surgical Intervention	A0705	Battery Problem	25/03/2021	25/03/2021
									A070504	Premature Discharge of Battery	25/03/2021	25/03/2021
10849	2021	All other reportable incidents	17882	Physiologic-monitoring defibrillation system	CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A07	Electrical /Electronic Property Problem		
							F27	Problem Identified During Non-Clinical Procedure	A07	Electrical /Electronic Property Problem		
10850	2021	All other reportable incidents	35223	Endocardial/interventricular septal pacing lead	CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A0401	Break		
									A070909	Oversensing		
							F1901	Additional Surgery	A0401	Break		
									A070909	Oversensing		
							F1905	Device Revision or Replacement	A0401	Break		
									A070909	Oversensing		
10851	2021	All other reportable incidents	35223	Endocardial/interventricular septal pacing lead	CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070909	Oversensing		
							F1901	Additional Surgery	A070909	Oversensing		
							F1905	Device Revision or Replacement	A070909	Oversensing		
10852	2021	All other reportable incidents	47265	Dual-chamber implantable pacemaker, rate-responsive	CT141	Electrodes/Leads and associated devices	E0112	Dizziness	A070101	Failure to Capture		
									A070909	Oversensing		
									A0712	Pacing Problem		
									A072201	High Impedance		
							E060101	Asystole	A070101	Failure to Capture		
									A070909	Oversensing		
									A0712	Pacing Problem		
									A072201	High Impedance		
							E060104	Bradycardia	A070101	Failure to Capture		
									A070909	Oversensing		
									A0712	Pacing Problem		
									A072201	High Impedance		
							E0717	Dyspnea	A070101	Failure to Capture		
									A070909	Oversensing		
									A0712	Pacing Problem		
									A072201	High Impedance		
							E2311	Discomfort	A070101	Failure to Capture		
									A070909	Oversensing		
									A0712	Pacing Problem		
									A072201	High Impedance		
							E2403	No Clinical Signs, Symptoms or Conditions	A070101	Failure to Capture		
									A070909	Oversensing		
									A0712	Pacing Problem		
									A072201	High Impedance		
							F1001	Absence of Treatment	A070101	Failure to Capture		
									A070909	Oversensing		
									A0712	Pacing Problem		
									A072201	High Impedance		
							F23	Unexpected Medical Intervention	A070101	Failure to Capture		
									A070909	Oversensing		
									A0712	Pacing Problem		
									A072201	High Impedance		
					CT145	Cardiovascular devices	E0112	Dizziness	A070101	Failure to Capture		
									A070909	Oversensing		
									A0712	Pacing Problem		
									A072201	High Impedance		
							E060101	Asystole	A070101	Failure to Capture		
									A070909	Oversensing		
									A0712	Pacing Problem		
									A072201	High Impedance		
							E060104	Bradycardia	A070101	Failure to Capture		
									A070909	Oversensing		
									A0712	Pacing Problem		
									A072201	High Impedance		
							E0717	Dyspnea	A070101	Failure to Capture		

									A070909	Oversensing		
									A0712	Pacing Problem		
									A072201	High Impedance		
								E2311	Discomfort	A070101	Failure to Capture		
										A070909	Oversensing		
										A0712	Pacing Problem		
										A072201	High Impedance		
								E2403	No Clinical Signs, Symptoms or Conditions	A070101	Failure to Capture		
										A070909	Oversensing		
										A0712	Pacing Problem		
										A072201	High Impedance		
								F1001	Absence of Treatment	A070101	Failure to Capture		
										A070909	Oversensing		
										A0712	Pacing Problem		
										A072201	High Impedance		
								F23	Unexpected Medical Intervention	A070101	Failure to Capture		
										A070909	Oversensing		
										A0712	Pacing Problem		
										A072201	High Impedance		
							CT1684	Cardiac pacemakers and associated devices	E0112	Dizziness	A070101	Failure to Capture		
											A070909	Oversensing		
											A0712	Pacing Problem		
											A072201	High Impedance		
									E060101	Asystole	A070101	Failure to Capture		
											A070909	Oversensing		
											A0712	Pacing Problem		
											A072201	High Impedance		
									E060104	Bradycardia	A070101	Failure to Capture		
											A070909	Oversensing		
											A0712	Pacing Problem		
											A072201	High Impedance		
									E0717	Dyspnea	A070101	Failure to Capture		
											A070909	Oversensing		
											A0712	Pacing Problem		
											A072201	High Impedance		
									E2311	Discomfort	A070101	Failure to Capture		
											A070909	Oversensing		
											A0712	Pacing Problem		
											A072201	High Impedance		
									E2403	No Clinical Signs, Symptoms or Conditions	A070101	Failure to Capture		
											A070909	Oversensing		
											A0712	Pacing Problem		
											A072201	High Impedance		
									F1001	Absence of Treatment	A070101	Failure to Capture		
											A070909	Oversensing		
											A0712	Pacing Problem		
											A072201	High Impedance		
									F23	Unexpected Medical Intervention	A070101	Failure to Capture		
											A070909	Oversensing		
											A0712	Pacing Problem		
											A072201	High Impedance		
							CT1685	Cardiac pacemakers	E0112	Dizziness	A070101	Failure to Capture		
											A070909	Oversensing		
											A0712	Pacing Problem		
											A072201	High Impedance		
									E060101	Asystole	A070101	Failure to Capture		
											A070909	Oversensing		
											A0712	Pacing Problem		
											A072201	High Impedance		
									E060104	Bradycardia	A070101	Failure to Capture		
											A070909	Oversensing		
											A0712	Pacing Problem		
											A072201	High Impedance		
									E0717	Dyspnea	A070101	Failure to Capture		
											A070909	Oversensing		
											A0712	Pacing Problem		
											A072201	High Impedance		
									E2311	Discomfort	A070101	Failure to Capture		
											A070909	Oversensing		
											A0712	Pacing Problem		
											A072201	High Impedance		
									E2403	No Clinical Signs, Symptoms or Conditions	A070101	Failure to Capture		
											A070909	Oversensing		
											A0712	Pacing Problem		
											A072201	High Impedance		
									F1001	Absence of Treatment	A070101	Failure to Capture		
											A070909	Oversensing		
											A0712	Pacing Problem		
											A072201	High Impedance		
									F23	Unexpected Medical Intervention	A070101	Failure to Capture		
											A070909	Oversensing		
											A0712	Pacing Problem		
											A072201	High Impedance		
							CT233	Surgical	E0112	Dizziness	A070101	Failure to Capture		
											A070909	Oversensing		
											A0712	Pacing Problem		
											A072201	High Impedance		
									E060101	Asystole	A070101	Failure to Capture		

									A070909	Oversensing		
									A0712	Pacing Problem		
									A072201	High Impedance		
								E060104	Bradycardia	A070101	Failure to Capture		
										A070909	Oversensing		
										A0712	Pacing Problem		
										A072201	High Impedance		
								E0717	Dyspnea	A070101	Failure to Capture		
										A070909	Oversensing		
										A0712	Pacing Problem		
										A072201	High Impedance		
								E2311	Discomfort	A070101	Failure to Capture		
										A070909	Oversensing		
										A0712	Pacing Problem		
										A072201	High Impedance		
								E2403	No Clinical Signs, Symptoms or Conditions	A070101	Failure to Capture		
										A070909	Oversensing		
										A0712	Pacing Problem		
										A072201	High Impedance		
								F1001	Absence of Treatment	A070101	Failure to Capture		
										A070909	Oversensing		
										A0712	Pacing Problem		
										A072201	High Impedance		
								F23	Unexpected Medical Intervention	A070101	Failure to Capture		
										A070909	Oversensing		
										A0712	Pacing Problem		
										A072201	High Impedance		
					CT2406	Implantable		E0112	Dizziness	A070101	Failure to Capture		
										A070909	Oversensing		
										A0712	Pacing Problem		
										A072201	High Impedance		
								E060101	Asystole	A070101	Failure to Capture		
										A070909	Oversensing		
										A0712	Pacing Problem		
										A072201	High Impedance		
								E060104	Bradycardia	A070101	Failure to Capture		
										A070909	Oversensing		
										A0712	Pacing Problem		
										A072201	High Impedance		
								E0717	Dyspnea	A070101	Failure to Capture		
										A070909	Oversensing		
										A0712	Pacing Problem		
										A072201	High Impedance		
								E2311	Discomfort	A070101	Failure to Capture		
										A070909	Oversensing		
										A0712	Pacing Problem		
										A072201	High Impedance		
								E2403	No Clinical Signs, Symptoms or Conditions	A070101	Failure to Capture		
										A070909	Oversensing		
										A0712	Pacing Problem		
										A072201	High Impedance		
								F1001	Absence of Treatment	A070101	Failure to Capture		
										A070909	Oversensing		
										A0712	Pacing Problem		
										A072201	High Impedance		
								F23	Unexpected Medical Intervention	A070101	Failure to Capture		
										A070909	Oversensing		
										A0712	Pacing Problem		
										A072201	High Impedance		
					CT2407	Active implantable		E0112	Dizziness	A070101	Failure to Capture		
										A070909	Oversensing		
										A0712	Pacing Problem		
										A072201	High Impedance		
								E060101	Asystole	A070101	Failure to Capture		
										A070909	Oversensing		
										A0712	Pacing Problem		
										A072201	High Impedance		
								E060104	Bradycardia	A070101	Failure to Capture		
										A070909	Oversensing		
										A0712	Pacing Problem		
										A072201	High Impedance		
								E0717	Dyspnea	A070101	Failure to Capture		
										A070909	Oversensing		
										A0712	Pacing Problem		
										A072201	High Impedance		
								E2311	Discomfort	A070101	Failure to Capture		
										A070909	Oversensing		
										A0712	Pacing Problem		
										A072201	High Impedance		
								E2403	No Clinical Signs, Symptoms or Conditions	A070101	Failure to Capture		
										A070909	Oversensing		
										A0712	Pacing Problem		
										A072201	High Impedance		
								F1001	Absence of Treatment	A070101	Failure to Capture		
										A070909	Oversensing		
										A0712	Pacing Problem		
										A072201	High Impedance		
								F23	Unexpected Medical Intervention	A070101	Failure to Capture		

									A070909	Oversensing		
									A0712	Pacing Problem		
									A072201	High Impedance		
					CT257	Implantable cardiac pacemakers	E0112	Dizziness	A070101	Failure to Capture		
									A070909	Oversensing		
									A0712	Pacing Problem		
									A072201	High Impedance		
							E060101	Asystole	A070101	Failure to Capture		
									A070909	Oversensing		
									A0712	Pacing Problem		
									A072201	High Impedance		
							E060104	Bradycardia	A070101	Failure to Capture		
									A070909	Oversensing		
									A0712	Pacing Problem		
									A072201	High Impedance		
							E0717	Dyspnea	A070101	Failure to Capture		
									A070909	Oversensing		
									A0712	Pacing Problem		
									A072201	High Impedance		
							E2311	Discomfort	A070101	Failure to Capture		
									A070909	Oversensing		
									A0712	Pacing Problem		
									A072201	High Impedance		
							E2403	No Clinical Signs, Symptoms or Conditions	A070101	Failure to Capture		
									A070909	Oversensing		
									A0712	Pacing Problem		
									A072201	High Impedance		
							F1001	Absence of Treatment	A070101	Failure to Capture		
									A070909	Oversensing		
									A0712	Pacing Problem		
									A072201	High Impedance		
							F23	Unexpected Medical Intervention	A070101	Failure to Capture		
									A070909	Oversensing		
									A0712	Pacing Problem		
									A072201	High Impedance		
					CT2727	Active implant devices/ancillaries	E0112	Dizziness	A070101	Failure to Capture		
									A070909	Oversensing		
									A0712	Pacing Problem		
									A072201	High Impedance		
							E060101	Asystole	A070101	Failure to Capture		
									A070909	Oversensing		
									A0712	Pacing Problem		
									A072201	High Impedance		
							E060104	Bradycardia	A070101	Failure to Capture		
									A070909	Oversensing		
									A0712	Pacing Problem		
									A072201	High Impedance		
							E0717	Dyspnea	A070101	Failure to Capture		
									A070909	Oversensing		
									A0712	Pacing Problem		
									A072201	High Impedance		
							E2311	Discomfort	A070101	Failure to Capture		
									A070909	Oversensing		
									A0712	Pacing Problem		
									A072201	High Impedance		
							E2403	No Clinical Signs, Symptoms or Conditions	A070101	Failure to Capture		
									A070909	Oversensing		
									A0712	Pacing Problem		
									A072201	High Impedance		
							F1001	Absence of Treatment	A070101	Failure to Capture		
									A070909	Oversensing		
									A0712	Pacing Problem		
									A072201	High Impedance		
							F23	Unexpected Medical Intervention	A070101	Failure to Capture		
									A070909	Oversensing		
									A0712	Pacing Problem		
									A072201	High Impedance		
					CT2750	Implantable cardiac electrical stimulators and associated devices	E0112	Dizziness	A070101	Failure to Capture		
									A070909	Oversensing		
									A0712	Pacing Problem		
									A072201	High Impedance		
							E060101	Asystole	A070101	Failure to Capture		
									A070909	Oversensing		
									A0712	Pacing Problem		
									A072201	High Impedance		
							E060104	Bradycardia	A070101	Failure to Capture		
									A070909	Oversensing		
									A0712	Pacing Problem		
									A072201	High Impedance		
							E0717	Dyspnea	A070101	Failure to Capture		
									A070909	Oversensing		
									A0712	Pacing Problem		
									A072201	High Impedance		
							E2311	Discomfort	A070101	Failure to Capture		
									A070909	Oversensing		
									A0712	Pacing Problem		
									A072201	High Impedance		
							E2403	No Clinical Signs, Symptoms or Conditions	A070101	Failure to Capture		

									A070909	Oversensing		
									A0712	Pacing Problem		
									A072201	High Impedance		
								F1001	Absence of Treatment	A070101	Failure to Capture		
										A070909	Oversensing		
										A0712	Pacing Problem		
										A072201	High Impedance		
								F23	Unexpected Medical Intervention	A070101	Failure to Capture		
										A070909	Oversensing		
										A0712	Pacing Problem		
										A072201	High Impedance		
						CT2866	Implantable cardiac electrical stimulators	E0112	Dizziness	A070101	Failure to Capture		
										A070909	Oversensing		
										A0712	Pacing Problem		
										A072201	High Impedance		
								E060101	Asystole	A070101	Failure to Capture		
										A070909	Oversensing		
										A0712	Pacing Problem		
										A072201	High Impedance		
								E060104	Bradycardia	A070101	Failure to Capture		
										A070909	Oversensing		
										A0712	Pacing Problem		
										A072201	High Impedance		
								E0717	Dyspnea	A070101	Failure to Capture		
										A070909	Oversensing		
										A0712	Pacing Problem		
										A072201	High Impedance		
								E2311	Discomfort	A070101	Failure to Capture		
										A070909	Oversensing		
										A0712	Pacing Problem		
										A072201	High Impedance		
								E2403	No Clinical Signs, Symptoms or Conditions	A070101	Failure to Capture		
										A070909	Oversensing		
										A0712	Pacing Problem		
										A072201	High Impedance		
								F1001	Absence of Treatment	A070101	Failure to Capture		
										A070909	Oversensing		
										A0712	Pacing Problem		
										A072201	High Impedance		
								F23	Unexpected Medical Intervention	A070101	Failure to Capture		
										A070909	Oversensing		
										A0712	Pacing Problem		
										A072201	High Impedance		
						CT2985	Active	E0112	Dizziness	A070101	Failure to Capture		
										A070909	Oversensing		
										A0712	Pacing Problem		
										A072201	High Impedance		
								E060101	Asystole	A070101	Failure to Capture		
										A070909	Oversensing		
										A0712	Pacing Problem		
										A072201	High Impedance		
								E060104	Bradycardia	A070101	Failure to Capture		
										A070909	Oversensing		
										A0712	Pacing Problem		
										A072201	High Impedance		
								E0717	Dyspnea	A070101	Failure to Capture		
										A070909	Oversensing		
										A0712	Pacing Problem		
										A072201	High Impedance		
								E2311	Discomfort	A070101	Failure to Capture		
										A070909	Oversensing		
										A0712	Pacing Problem		
										A072201	High Impedance		
								E2403	No Clinical Signs, Symptoms or Conditions	A070101	Failure to Capture		
										A070909	Oversensing		
										A0712	Pacing Problem		
										A072201	High Impedance		
								F1001	Absence of Treatment	A070101	Failure to Capture		
										A070909	Oversensing		
										A0712	Pacing Problem		
										A072201	High Impedance		
								F23	Unexpected Medical Intervention	A070101	Failure to Capture		
										A070909	Oversensing		
										A0712	Pacing Problem		
										A072201	High Impedance		
						CT3001	Implantable metal	E0112	Dizziness	A070101	Failure to Capture		
										A070909	Oversensing		
										A0712	Pacing Problem		
										A072201	High Impedance		
								E060101	Asystole	A070101	Failure to Capture		
										A070909	Oversensing		
										A0712	Pacing Problem		
										A072201	High Impedance		
								E060104	Bradycardia	A070101	Failure to Capture		
										A070909	Oversensing		
										A0712	Pacing Problem		
										A072201	High Impedance		
								E0717	Dyspnea	A070101	Failure to Capture		

									A070909	Oversensing		
									A0712	Pacing Problem		
									A072201	High Impedance		
								E2311	Discomfort	A070101	Failure to Capture		
										A070909	Oversensing		
										A0712	Pacing Problem		
										A072201	High Impedance		
								E2403	No Clinical Signs, Symptoms or Conditions	A070101	Failure to Capture		
										A070909	Oversensing		
										A0712	Pacing Problem		
										A072201	High Impedance		
								F1001	Absence of Treatment	A070101	Failure to Capture		
										A070909	Oversensing		
										A0712	Pacing Problem		
										A072201	High Impedance		
								F23	Unexpected Medical Intervention	A070101	Failure to Capture		
										A070909	Oversensing		
										A0712	Pacing Problem		
										A072201	High Impedance		
							CT3002	Implantable synthetic polymer	E0112	Dizziness	A070101	Failure to Capture		
											A070909	Oversensing		
											A0712	Pacing Problem		
											A072201	High Impedance		
									E060101	Asystole	A070101	Failure to Capture		
											A070909	Oversensing		
											A0712	Pacing Problem		
											A072201	High Impedance		
									E060104	Bradycardia	A070101	Failure to Capture		
											A070909	Oversensing		
											A0712	Pacing Problem		
											A072201	High Impedance		
									E0717	Dyspnea	A070101	Failure to Capture		
											A070909	Oversensing		
											A0712	Pacing Problem		
											A072201	High Impedance		
									E2311	Discomfort	A070101	Failure to Capture		
											A070909	Oversensing		
											A0712	Pacing Problem		
											A072201	High Impedance		
									E2403	No Clinical Signs, Symptoms or Conditions	A070101	Failure to Capture		
											A070909	Oversensing		
											A0712	Pacing Problem		
											A072201	High Impedance		
									F1001	Absence of Treatment	A070101	Failure to Capture		
											A070909	Oversensing		
											A0712	Pacing Problem		
											A072201	High Impedance		
									F23	Unexpected Medical Intervention	A070101	Failure to Capture		
											A070909	Oversensing		
											A0712	Pacing Problem		
											A072201	High Impedance		
							CT3006	Implantable inorganic material	E0112	Dizziness	A070101	Failure to Capture		
											A070909	Oversensing		
											A0712	Pacing Problem		
											A072201	High Impedance		
									E060101	Asystole	A070101	Failure to Capture		
											A070909	Oversensing		
											A0712	Pacing Problem		
											A072201	High Impedance		
									E060104	Bradycardia	A070101	Failure to Capture		
											A070909	Oversensing		
											A0712	Pacing Problem		
											A072201	High Impedance		
									E0717	Dyspnea	A070101	Failure to Capture		
											A070909	Oversensing		
											A0712	Pacing Problem		
											A072201	High Impedance		
									E2311	Discomfort	A070101	Failure to Capture		
											A070909	Oversensing		
											A0712	Pacing Problem		
											A072201	High Impedance		
									E2403	No Clinical Signs, Symptoms or Conditions	A070101	Failure to Capture		
											A070909	Oversensing		
											A0712	Pacing Problem		
											A072201	High Impedance		
									F1001	Absence of Treatment	A070101	Failure to Capture		
											A070909	Oversensing		
											A0712	Pacing Problem		
											A072201	High Impedance		
									F23	Unexpected Medical Intervention	A070101	Failure to Capture		
											A070909	Oversensing		
											A0712	Pacing Problem		
											A072201	High Impedance		
							CT314	Electrically powered	E0112	Dizziness	A070101	Failure to Capture		
											A070909	Oversensing		
											A0712	Pacing Problem		
											A072201	High Impedance		
									E060101	Asystole	A070101	Failure to Capture		

									A070909	Oversensing		
									A0712	Pacing Problem		
									A072201	High Impedance		
								E060104	Bradycardia	A070101	Failure to Capture		
										A070909	Oversensing		
										A0712	Pacing Problem		
										A072201	High Impedance		
								E0717	Dyspnea	A070101	Failure to Capture		
										A070909	Oversensing		
										A0712	Pacing Problem		
										A072201	High Impedance		
								E2311	Discomfort	A070101	Failure to Capture		
										A070909	Oversensing		
										A0712	Pacing Problem		
										A072201	High Impedance		
								E2403	No Clinical Signs, Symptoms or Conditions	A070101	Failure to Capture		
										A070909	Oversensing		
										A0712	Pacing Problem		
										A072201	High Impedance		
								F1001	Absence of Treatment	A070101	Failure to Capture		
										A070909	Oversensing		
										A0712	Pacing Problem		
										A072201	High Impedance		
								F23	Unexpected Medical Intervention	A070101	Failure to Capture		
										A070909	Oversensing		
										A0712	Pacing Problem		
										A072201	High Impedance		
						CT321	Long-term surgical invasive	E0112	Dizziness	A070101	Failure to Capture		
										A070909	Oversensing		
										A0712	Pacing Problem		
										A072201	High Impedance		
								E060101	Asystole	A070101	Failure to Capture		
										A070909	Oversensing		
										A0712	Pacing Problem		
										A072201	High Impedance		
								E060104	Bradycardia	A070101	Failure to Capture		
										A070909	Oversensing		
										A0712	Pacing Problem		
										A072201	High Impedance		
								E0717	Dyspnea	A070101	Failure to Capture		
										A070909	Oversensing		
										A0712	Pacing Problem		
										A072201	High Impedance		
								E2311	Discomfort	A070101	Failure to Capture		
										A070909	Oversensing		
										A0712	Pacing Problem		
										A072201	High Impedance		
								E2403	No Clinical Signs, Symptoms or Conditions	A070101	Failure to Capture		
										A070909	Oversensing		
										A0712	Pacing Problem		
										A072201	High Impedance		
								F1001	Absence of Treatment	A070101	Failure to Capture		
										A070909	Oversensing		
										A0712	Pacing Problem		
										A072201	High Impedance		
								F23	Unexpected Medical Intervention	A070101	Failure to Capture		
										A070909	Oversensing		
										A0712	Pacing Problem		
										A072201	High Impedance		
						CT449	Generators and associated devices	E0112	Dizziness	A070101	Failure to Capture		
										A070909	Oversensing		
										A0712	Pacing Problem		
										A072201	High Impedance		
								E060101	Asystole	A070101	Failure to Capture		
										A070909	Oversensing		
										A0712	Pacing Problem		
										A072201	High Impedance		
								E060104	Bradycardia	A070101	Failure to Capture		
										A070909	Oversensing		
										A0712	Pacing Problem		
										A072201	High Impedance		
								E0717	Dyspnea	A070101	Failure to Capture		
										A070909	Oversensing		
										A0712	Pacing Problem		
										A072201	High Impedance		
								E2311	Discomfort	A070101	Failure to Capture		
										A070909	Oversensing		
										A0712	Pacing Problem		
										A072201	High Impedance		
								E2403	No Clinical Signs, Symptoms or Conditions	A070101	Failure to Capture		
										A070909	Oversensing		
										A0712	Pacing Problem		
										A072201	High Impedance		
								F1001	Absence of Treatment	A070101	Failure to Capture		
										A070909	Oversensing		
										A0712	Pacing Problem		
										A072201	High Impedance		
								F23	Unexpected Medical Intervention	A070101	Failure to Capture		

									A070909	Oversensing		
									A0712	Pacing Problem		
									A072201	High Impedance		
					CT981	Single use	E0112	Dizziness	A070101	Failure to Capture		
									A070909	Oversensing		
									A0712	Pacing Problem		
									A072201	High Impedance		
							E060101	Asystole	A070101	Failure to Capture		
									A070909	Oversensing		
									A0712	Pacing Problem		
									A072201	High Impedance		
							E060104	Bradycardia	A070101	Failure to Capture		
									A070909	Oversensing		
									A0712	Pacing Problem		
									A072201	High Impedance		
							E0717	Dyspnea	A070101	Failure to Capture		
									A070909	Oversensing		
									A0712	Pacing Problem		
									A072201	High Impedance		
							E2311	Discomfort	A070101	Failure to Capture		
									A070909	Oversensing		
									A0712	Pacing Problem		
									A072201	High Impedance		
							E2403	No Clinical Signs, Symptoms or Conditions	A070101	Failure to Capture		
									A070909	Oversensing		
									A0712	Pacing Problem		
									A072201	High Impedance		
							F1001	Absence of Treatment	A070101	Failure to Capture		
									A070909	Oversensing		
									A0712	Pacing Problem		
									A072201	High Impedance		
							F23	Unexpected Medical Intervention	A070101	Failure to Capture		
									A070909	Oversensing		
									A0712	Pacing Problem		
									A072201	High Impedance		
					CT983	Surgical invasive	E0112	Dizziness	A070101	Failure to Capture		
									A070909	Oversensing		
									A0712	Pacing Problem		
									A072201	High Impedance		
							E060101	Asystole	A070101	Failure to Capture		
									A070909	Oversensing		
									A0712	Pacing Problem		
									A072201	High Impedance		
							E060104	Bradycardia	A070101	Failure to Capture		
									A070909	Oversensing		
									A0712	Pacing Problem		
									A072201	High Impedance		
							E0717	Dyspnea	A070101	Failure to Capture		
									A070909	Oversensing		
									A0712	Pacing Problem		
									A072201	High Impedance		
							E2311	Discomfort	A070101	Failure to Capture		
									A070909	Oversensing		
									A0712	Pacing Problem		
									A072201	High Impedance		
							E2403	No Clinical Signs, Symptoms or Conditions	A070101	Failure to Capture		
									A070909	Oversensing		
									A0712	Pacing Problem		
									A072201	High Impedance		
							F1001	Absence of Treatment	A070101	Failure to Capture		
									A070909	Oversensing		
									A0712	Pacing Problem		
									A072201	High Impedance		
							F23	Unexpected Medical Intervention	A070101	Failure to Capture		
									A070909	Oversensing		
									A0712	Pacing Problem		
									A072201	High Impedance		
10853	2021	All other reportable incidents	35852	Single-chamber implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem	24/06/2021	24/06/2021
									A070504	Premature Discharge of Battery	24/06/2021	24/06/2021
							F1905	Device Revision or Replacement	A0705	Battery Problem	24/06/2021	24/06/2021
									A070504	Premature Discharge of Battery	24/06/2021	24/06/2021
10854	2021	All other reportable incidents	35852	Single-chamber implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem	20/04/2021	20/04/2021
									A070504	Premature Discharge of Battery	20/04/2021	20/04/2021
							F19	Surgical Intervention	A0705	Battery Problem	20/04/2021	20/04/2021
									A070504	Premature Discharge of Battery	20/04/2021	20/04/2021
10855	2021	All other reportable incidents	35852	Single-chamber implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	09/06/2021	09/06/2021
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	09/06/2021	09/06/2021
							F19	Surgical Intervention	A070504	Premature Discharge of Battery	09/06/2021	09/06/2021
10856	2021	All other reportable incidents	35852	Single-chamber implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A071201	Failure to Convert Rhythm		
									A23	Use of Device Problem		
							F26	No Health Consequences or Impact	A071201	Failure to Convert Rhythm		
									A23	Use of Device Problem		
10857	2021	All other reportable incidents	47263	Cardiac resynchronization therapy implantable pacemaker	CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A0908	Incorrect, Inadequate or Imprecise Result or Readings	25/03/2021	25/03/2021
							F08	Hospitalization or Prolonged Hospitalization	A0908	Incorrect, Inadequate or Imprecise Result or Readings	25/03/2021	25/03/2021
							F1905	Device Revision or Replacement	A0908	Incorrect, Inadequate or Imprecise Result or Readings	25/03/2021	25/03/2021

					CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A0908	Incorrect, Inadequate or Imprecise Result or Readings	25/03/2021	25/03/2021
							F08	Hospitalization or Prolonged Hospitalization	A0908	Incorrect, Inadequate or Imprecise Result or Readings	25/03/2021	25/03/2021
							F1905	Device Revision or Replacement	A0908	Incorrect, Inadequate or Imprecise Result or Readings	25/03/2021	25/03/2021
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A0908	Incorrect, Inadequate or Imprecise Result or Readings	25/03/2021	25/03/2021
							F08	Hospitalization or Prolonged Hospitalization	A0908	Incorrect, Inadequate or Imprecise Result or Readings	25/03/2021	25/03/2021
							F1905	Device Revision or Replacement	A0908	Incorrect, Inadequate or Imprecise Result or Readings	25/03/2021	25/03/2021
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A0908	Incorrect, Inadequate or Imprecise Result or Readings	25/03/2021	25/03/2021
							F08	Hospitalization or Prolonged Hospitalization	A0908	Incorrect, Inadequate or Imprecise Result or Readings	25/03/2021	25/03/2021
							F1905	Device Revision or Replacement	A0908	Incorrect, Inadequate or Imprecise Result or Readings	25/03/2021	25/03/2021
10858	2021	Unanticipated serious deterioration in state of health	47265	Dual-chamber implantable pacemaker, rate-responsive	CT141	Electrodes/Leads and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	27/07/2021	27/07/2021
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	27/07/2021	27/07/2021
							F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A070504	Premature Discharge of Battery	27/07/2021	27/07/2021
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	27/07/2021	27/07/2021
					CT145	Cardiovascular devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	27/07/2021	27/07/2021
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	27/07/2021	27/07/2021
							F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A070504	Premature Discharge of Battery	27/07/2021	27/07/2021
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	27/07/2021	27/07/2021
					CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	27/07/2021	27/07/2021
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	27/07/2021	27/07/2021
							F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A070504	Premature Discharge of Battery	27/07/2021	27/07/2021
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	27/07/2021	27/07/2021
					CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	27/07/2021	27/07/2021
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	27/07/2021	27/07/2021
							F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A070504	Premature Discharge of Battery	27/07/2021	27/07/2021
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	27/07/2021	27/07/2021
					CT233	Surgical	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	27/07/2021	27/07/2021
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	27/07/2021	27/07/2021
							F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A070504	Premature Discharge of Battery	27/07/2021	27/07/2021
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	27/07/2021	27/07/2021
					CT2406	Implantable	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	27/07/2021	27/07/2021
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	27/07/2021	27/07/2021
							F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A070504	Premature Discharge of Battery	27/07/2021	27/07/2021
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	27/07/2021	27/07/2021
					CT2407	Active implantable	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	27/07/2021	27/07/2021
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	27/07/2021	27/07/2021
							F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A070504	Premature Discharge of Battery	27/07/2021	27/07/2021
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	27/07/2021	27/07/2021
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	27/07/2021	27/07/2021
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	27/07/2021	27/07/2021
							F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A070504	Premature Discharge of Battery	27/07/2021	27/07/2021
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	27/07/2021	27/07/2021
					CT2727	Active implant devices/ancillaries	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	27/07/2021	27/07/2021
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	27/07/2021	27/07/2021
							F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A070504	Premature Discharge of Battery	27/07/2021	27/07/2021
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	27/07/2021	27/07/2021
					CT2750	Implantable cardiac electrical stimulators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	27/07/2021	27/07/2021
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	27/07/2021	27/07/2021
							F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A070504	Premature Discharge of Battery	27/07/2021	27/07/2021
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	27/07/2021	27/07/2021
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	27/07/2021	27/07/2021
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	27/07/2021	27/07/2021
							F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A070504	Premature Discharge of Battery	27/07/2021	27/07/2021
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	27/07/2021	27/07/2021
					CT2985	Active	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	27/07/2021	27/07/2021
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	27/07/2021	27/07/2021
							F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A070504	Premature Discharge of Battery	27/07/2021	27/07/2021
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	27/07/2021	27/07/2021
					CT3001	Implantable metal	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	27/07/2021	27/07/2021
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	27/07/2021	27/07/2021
							F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A070504	Premature Discharge of Battery	27/07/2021	27/07/2021
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	27/07/2021	27/07/2021
					CT3002	Implantable synthetic polymer	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	27/07/2021	27/07/2021
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	27/07/2021	27/07/2021
							F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A070504	Premature Discharge of Battery	27/07/2021	27/07/2021
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	27/07/2021	27/07/2021
					CT3006	Implantable inorganic material	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	27/07/2021	27/07/2021
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	27/07/2021	27/07/2021
							F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A070504	Premature Discharge of Battery	27/07/2021	27/07/2021
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	27/07/2021	27/07/2021
					CT314	Electrically powered	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	27/07/2021	27/07/2021
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	27/07/2021	27/07/2021
							F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A070504	Premature Discharge of Battery	27/07/2021	27/07/2021
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	27/07/2021	27/07/2021
					CT321	Long-term surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	27/07/2021	27/07/2021
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	27/07/2021	27/07/2021
							F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A070504	Premature Discharge of Battery	27/07/2021	27/07/2021
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	27/07/2021	27/07/2021
					CT449	Generators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	27/07/2021	27/07/2021
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	27/07/2021	27/07/2021
							F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A070504	Premature Discharge of Battery	27/07/2021	27/07/2021
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	27/07/2021	27/07/2021

					CT981	Single use	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	27/07/2021	27/07/2021
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	27/07/2021	27/07/2021
							F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A070504	Premature Discharge of Battery	27/07/2021	27/07/2021
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	27/07/2021	27/07/2021
					CT983	Surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	27/07/2021	27/07/2021
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	27/07/2021	27/07/2021
							F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A070504	Premature Discharge of Battery	27/07/2021	27/07/2021
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	27/07/2021	27/07/2021
10859	2021	All other reportable incidents	36241	Pacemaker/defibrillator lead	CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A05	Mechanical Problem	18/03/2021	18/03/2021
									A070102	High Capture Threshold	18/03/2021	18/03/2021
							F1905	Device Revision or Replacement	A05	Mechanical Problem	18/03/2021	18/03/2021
									A070102	High Capture Threshold	18/03/2021	18/03/2021
							F1908	Prolonged Surgery	A05	Mechanical Problem	18/03/2021	18/03/2021
									A070102	High Capture Threshold	18/03/2021	18/03/2021
10860	2021	All other reportable incidents	47265	Dual-chamber implantable pacemaker, rate-responsive	CT141	Electrodes/Leads and associated devices	E060104	Bradycardia	A0712	Pacing Problem	19/03/2021	19/03/2021
									A0908	Incorrect, Inadequate or Imprecise Result or Readings	19/03/2021	19/03/2021
							E2403	No Clinical Signs, Symptoms or Conditions	A0712	Pacing Problem	19/03/2021	19/03/2021
									A0908	Incorrect, Inadequate or Imprecise Result or Readings	19/03/2021	19/03/2021
							F08	Hospitalization or Prolonged Hospitalization	A0712	Pacing Problem	19/03/2021	19/03/2021
									A0908	Incorrect, Inadequate or Imprecise Result or Readings	19/03/2021	19/03/2021
							F12	Serious Injury/ Illness/ Impairment	A0712	Pacing Problem	19/03/2021	19/03/2021
									A0908	Incorrect, Inadequate or Imprecise Result or Readings	19/03/2021	19/03/2021
							F1905	Device Revision or Replacement	A0712	Pacing Problem	19/03/2021	19/03/2021
									A0908	Incorrect, Inadequate or Imprecise Result or Readings	19/03/2021	19/03/2021
					CT145	Cardiovascular devices	E060104	Bradycardia	A0712	Pacing Problem	19/03/2021	19/03/2021
									A0908	Incorrect, Inadequate or Imprecise Result or Readings	19/03/2021	19/03/2021
							E2403	No Clinical Signs, Symptoms or Conditions	A0712	Pacing Problem	19/03/2021	19/03/2021
									A0908	Incorrect, Inadequate or Imprecise Result or Readings	19/03/2021	19/03/2021
							F08	Hospitalization or Prolonged Hospitalization	A0712	Pacing Problem	19/03/2021	19/03/2021
									A0908	Incorrect, Inadequate or Imprecise Result or Readings	19/03/2021	19/03/2021
							F12	Serious Injury/ Illness/ Impairment	A0712	Pacing Problem	19/03/2021	19/03/2021
									A0908	Incorrect, Inadequate or Imprecise Result or Readings	19/03/2021	19/03/2021
							F1905	Device Revision or Replacement	A0712	Pacing Problem	19/03/2021	19/03/2021
									A0908	Incorrect, Inadequate or Imprecise Result or Readings	19/03/2021	19/03/2021
					CT1684	Cardiac pacemakers and associated devices	E060104	Bradycardia	A0712	Pacing Problem	19/03/2021	19/03/2021
									A0908	Incorrect, Inadequate or Imprecise Result or Readings	19/03/2021	19/03/2021
							E2403	No Clinical Signs, Symptoms or Conditions	A0712	Pacing Problem	19/03/2021	19/03/2021
									A0908	Incorrect, Inadequate or Imprecise Result or Readings	19/03/2021	19/03/2021
							F08	Hospitalization or Prolonged Hospitalization	A0712	Pacing Problem	19/03/2021	19/03/2021
									A0908	Incorrect, Inadequate or Imprecise Result or Readings	19/03/2021	19/03/2021
							F12	Serious Injury/ Illness/ Impairment	A0712	Pacing Problem	19/03/2021	19/03/2021
									A0908	Incorrect, Inadequate or Imprecise Result or Readings	19/03/2021	19/03/2021
							F1905	Device Revision or Replacement	A0712	Pacing Problem	19/03/2021	19/03/2021
									A0908	Incorrect, Inadequate or Imprecise Result or Readings	19/03/2021	19/03/2021
					CT1685	Cardiac pacemakers	E060104	Bradycardia	A0712	Pacing Problem	19/03/2021	19/03/2021
									A0908	Incorrect, Inadequate or Imprecise Result or Readings	19/03/2021	19/03/2021
							E2403	No Clinical Signs, Symptoms or Conditions	A0712	Pacing Problem	19/03/2021	19/03/2021
									A0908	Incorrect, Inadequate or Imprecise Result or Readings	19/03/2021	19/03/2021
							F08	Hospitalization or Prolonged Hospitalization	A0712	Pacing Problem	19/03/2021	19/03/2021
									A0908	Incorrect, Inadequate or Imprecise Result or Readings	19/03/2021	19/03/2021
							F12	Serious Injury/ Illness/ Impairment	A0712	Pacing Problem	19/03/2021	19/03/2021
									A0908	Incorrect, Inadequate or Imprecise Result or Readings	19/03/2021	19/03/2021
							F1905	Device Revision or Replacement	A0712	Pacing Problem	19/03/2021	19/03/2021
									A0908	Incorrect, Inadequate or Imprecise Result or Readings	19/03/2021	19/03/2021
					CT233	Surgical	E060104	Bradycardia	A0712	Pacing Problem	19/03/2021	19/03/2021
									A0908	Incorrect, Inadequate or Imprecise Result or Readings	19/03/2021	19/03/2021
							E2403	No Clinical Signs, Symptoms or Conditions	A0712	Pacing Problem	19/03/2021	19/03/2021
									A0908	Incorrect, Inadequate or Imprecise Result or Readings	19/03/2021	19/03/2021
							F08	Hospitalization or Prolonged Hospitalization	A0712	Pacing Problem	19/03/2021	19/03/2021
									A0908	Incorrect, Inadequate or Imprecise Result or Readings	19/03/2021	19/03/2021
							F12	Serious Injury/ Illness/ Impairment	A0712	Pacing Problem	19/03/2021	19/03/2021
									A0908	Incorrect, Inadequate or Imprecise Result or Readings	19/03/2021	19/03/2021
							F1905	Device Revision or Replacement	A0712	Pacing Problem	19/03/2021	19/03/2021
									A0908	Incorrect, Inadequate or Imprecise Result or Readings	19/03/2021	19/03/2021
					CT2406	Implantable	E060104	Bradycardia	A0712	Pacing Problem	19/03/2021	19/03/2021
									A0908	Incorrect, Inadequate or Imprecise Result or Readings	19/03/2021	19/03/2021
							E2403	No Clinical Signs, Symptoms or Conditions	A0712	Pacing Problem	19/03/2021	19/03/2021

									A0908	Incorrect, Inadequate or Imprecise Result or Readings	19/03/2021	19/03/2021
							F08	Hospitalization or Prolonged Hospitalization	A0712	Pacing Problem	19/03/2021	19/03/2021
									A0908	Incorrect, Inadequate or Imprecise Result or Readings	19/03/2021	19/03/2021
							F12	Serious Injury/ Illness/ Impairment	A0712	Pacing Problem	19/03/2021	19/03/2021
									A0908	Incorrect, Inadequate or Imprecise Result or Readings	19/03/2021	19/03/2021
							F1905	Device Revision or Replacement	A0712	Pacing Problem	19/03/2021	19/03/2021
									A0908	Incorrect, Inadequate or Imprecise Result or Readings	19/03/2021	19/03/2021
					CT2407	Active implantable	E060104	Bradycardia	A0712	Pacing Problem	19/03/2021	19/03/2021
									A0908	Incorrect, Inadequate or Imprecise Result or Readings	19/03/2021	19/03/2021
							E2403	No Clinical Signs, Symptoms or Conditions	A0712	Pacing Problem	19/03/2021	19/03/2021
									A0908	Incorrect, Inadequate or Imprecise Result or Readings	19/03/2021	19/03/2021
							F08	Hospitalization or Prolonged Hospitalization	A0712	Pacing Problem	19/03/2021	19/03/2021
									A0908	Incorrect, Inadequate or Imprecise Result or Readings	19/03/2021	19/03/2021
							F12	Serious Injury/ Illness/ Impairment	A0712	Pacing Problem	19/03/2021	19/03/2021
									A0908	Incorrect, Inadequate or Imprecise Result or Readings	19/03/2021	19/03/2021
							F1905	Device Revision or Replacement	A0712	Pacing Problem	19/03/2021	19/03/2021
									A0908	Incorrect, Inadequate or Imprecise Result or Readings	19/03/2021	19/03/2021
					CT257	Implantable cardiac pacemakers	E060104	Bradycardia	A0712	Pacing Problem	19/03/2021	19/03/2021
									A0908	Incorrect, Inadequate or Imprecise Result or Readings	19/03/2021	19/03/2021
							E2403	No Clinical Signs, Symptoms or Conditions	A0712	Pacing Problem	19/03/2021	19/03/2021
									A0908	Incorrect, Inadequate or Imprecise Result or Readings	19/03/2021	19/03/2021
							F08	Hospitalization or Prolonged Hospitalization	A0712	Pacing Problem	19/03/2021	19/03/2021
									A0908	Incorrect, Inadequate or Imprecise Result or Readings	19/03/2021	19/03/2021
							F12	Serious Injury/ Illness/ Impairment	A0712	Pacing Problem	19/03/2021	19/03/2021
									A0908	Incorrect, Inadequate or Imprecise Result or Readings	19/03/2021	19/03/2021
							F1905	Device Revision or Replacement	A0712	Pacing Problem	19/03/2021	19/03/2021
									A0908	Incorrect, Inadequate or Imprecise Result or Readings	19/03/2021	19/03/2021
					CT2727	Active implant devices/ancillaries	E060104	Bradycardia	A0712	Pacing Problem	19/03/2021	19/03/2021
									A0908	Incorrect, Inadequate or Imprecise Result or Readings	19/03/2021	19/03/2021
							E2403	No Clinical Signs, Symptoms or Conditions	A0712	Pacing Problem	19/03/2021	19/03/2021
									A0908	Incorrect, Inadequate or Imprecise Result or Readings	19/03/2021	19/03/2021
							F08	Hospitalization or Prolonged Hospitalization	A0712	Pacing Problem	19/03/2021	19/03/2021
									A0908	Incorrect, Inadequate or Imprecise Result or Readings	19/03/2021	19/03/2021
							F12	Serious Injury/ Illness/ Impairment	A0712	Pacing Problem	19/03/2021	19/03/2021
									A0908	Incorrect, Inadequate or Imprecise Result or Readings	19/03/2021	19/03/2021
							F1905	Device Revision or Replacement	A0712	Pacing Problem	19/03/2021	19/03/2021
									A0908	Incorrect, Inadequate or Imprecise Result or Readings	19/03/2021	19/03/2021
					CT2750	Implantable cardiac electrical stimulators and associated devices	E060104	Bradycardia	A0712	Pacing Problem	19/03/2021	19/03/2021
									A0908	Incorrect, Inadequate or Imprecise Result or Readings	19/03/2021	19/03/2021
							E2403	No Clinical Signs, Symptoms or Conditions	A0712	Pacing Problem	19/03/2021	19/03/2021
									A0908	Incorrect, Inadequate or Imprecise Result or Readings	19/03/2021	19/03/2021
							F08	Hospitalization or Prolonged Hospitalization	A0712	Pacing Problem	19/03/2021	19/03/2021
									A0908	Incorrect, Inadequate or Imprecise Result or Readings	19/03/2021	19/03/2021
							F12	Serious Injury/ Illness/ Impairment	A0712	Pacing Problem	19/03/2021	19/03/2021
									A0908	Incorrect, Inadequate or Imprecise Result or Readings	19/03/2021	19/03/2021
							F1905	Device Revision or Replacement	A0712	Pacing Problem	19/03/2021	19/03/2021
									A0908	Incorrect, Inadequate or Imprecise Result or Readings	19/03/2021	19/03/2021
					CT2866	Implantable cardiac electrical stimulators	E060104	Bradycardia	A0712	Pacing Problem	19/03/2021	19/03/2021
									A0908	Incorrect, Inadequate or Imprecise Result or Readings	19/03/2021	19/03/2021
							E2403	No Clinical Signs, Symptoms or Conditions	A0712	Pacing Problem	19/03/2021	19/03/2021
									A0908	Incorrect, Inadequate or Imprecise Result or Readings	19/03/2021	19/03/2021
							F08	Hospitalization or Prolonged Hospitalization	A0712	Pacing Problem	19/03/2021	19/03/2021
									A0908	Incorrect, Inadequate or Imprecise Result or Readings	19/03/2021	19/03/2021
							F12	Serious Injury/ Illness/ Impairment	A0712	Pacing Problem	19/03/2021	19/03/2021
									A0908	Incorrect, Inadequate or Imprecise Result or Readings	19/03/2021	19/03/2021
							F1905	Device Revision or Replacement	A0712	Pacing Problem	19/03/2021	19/03/2021
									A0908	Incorrect, Inadequate or Imprecise Result or Readings	19/03/2021	19/03/2021
					CT2985	Active	E060104	Bradycardia	A0712	Pacing Problem	19/03/2021	19/03/2021
									A0908	Incorrect, Inadequate or Imprecise Result or Readings	19/03/2021	19/03/2021
							E2403	No Clinical Signs, Symptoms or Conditions	A0712	Pacing Problem	19/03/2021	19/03/2021
									A0908	Incorrect, Inadequate or Imprecise Result or Readings	19/03/2021	19/03/2021
							F08	Hospitalization or Prolonged Hospitalization	A0712	Pacing Problem	19/03/2021	19/03/2021
									A0908	Incorrect, Inadequate or Imprecise Result or Readings	19/03/2021	19/03/2021

							F12	Serious Injury/ Illness/ Impairment	A0712	Pacing Problem	19/03/2021	19/03/2021
									A0908	Incorrect, Inadequate or Imprecise Result or Readings	19/03/2021	19/03/2021
							F1905	Device Revision or Replacement	A0712	Pacing Problem	19/03/2021	19/03/2021
									A0908	Incorrect, Inadequate or Imprecise Result or Readings	19/03/2021	19/03/2021
					CT3001	Implantable metal	E060104	Bradycardia	A0712	Pacing Problem	19/03/2021	19/03/2021
									A0908	Incorrect, Inadequate or Imprecise Result or Readings	19/03/2021	19/03/2021
							E2403	No Clinical Signs, Symptoms or Conditions	A0712	Pacing Problem	19/03/2021	19/03/2021
									A0908	Incorrect, Inadequate or Imprecise Result or Readings	19/03/2021	19/03/2021
							F08	Hospitalization or Prolonged Hospitalization	A0712	Pacing Problem	19/03/2021	19/03/2021
									A0908	Incorrect, Inadequate or Imprecise Result or Readings	19/03/2021	19/03/2021
							F12	Serious Injury/ Illness/ Impairment	A0712	Pacing Problem	19/03/2021	19/03/2021
									A0908	Incorrect, Inadequate or Imprecise Result or Readings	19/03/2021	19/03/2021
							F1905	Device Revision or Replacement	A0712	Pacing Problem	19/03/2021	19/03/2021
									A0908	Incorrect, Inadequate or Imprecise Result or Readings	19/03/2021	19/03/2021
					CT3002	Implantable synthetic polymer	E060104	Bradycardia	A0712	Pacing Problem	19/03/2021	19/03/2021
									A0908	Incorrect, Inadequate or Imprecise Result or Readings	19/03/2021	19/03/2021
							E2403	No Clinical Signs, Symptoms or Conditions	A0712	Pacing Problem	19/03/2021	19/03/2021
									A0908	Incorrect, Inadequate or Imprecise Result or Readings	19/03/2021	19/03/2021
							F08	Hospitalization or Prolonged Hospitalization	A0712	Pacing Problem	19/03/2021	19/03/2021
									A0908	Incorrect, Inadequate or Imprecise Result or Readings	19/03/2021	19/03/2021
							F12	Serious Injury/ Illness/ Impairment	A0712	Pacing Problem	19/03/2021	19/03/2021
									A0908	Incorrect, Inadequate or Imprecise Result or Readings	19/03/2021	19/03/2021
							F1905	Device Revision or Replacement	A0712	Pacing Problem	19/03/2021	19/03/2021
									A0908	Incorrect, Inadequate or Imprecise Result or Readings	19/03/2021	19/03/2021
					CT3006	Implantable inorganic material	E060104	Bradycardia	A0712	Pacing Problem	19/03/2021	19/03/2021
									A0908	Incorrect, Inadequate or Imprecise Result or Readings	19/03/2021	19/03/2021
							E2403	No Clinical Signs, Symptoms or Conditions	A0712	Pacing Problem	19/03/2021	19/03/2021
									A0908	Incorrect, Inadequate or Imprecise Result or Readings	19/03/2021	19/03/2021
							F08	Hospitalization or Prolonged Hospitalization	A0712	Pacing Problem	19/03/2021	19/03/2021
									A0908	Incorrect, Inadequate or Imprecise Result or Readings	19/03/2021	19/03/2021
							F12	Serious Injury/ Illness/ Impairment	A0712	Pacing Problem	19/03/2021	19/03/2021
									A0908	Incorrect, Inadequate or Imprecise Result or Readings	19/03/2021	19/03/2021
							F1905	Device Revision or Replacement	A0712	Pacing Problem	19/03/2021	19/03/2021
									A0908	Incorrect, Inadequate or Imprecise Result or Readings	19/03/2021	19/03/2021
					CT314	Electrically powered	E060104	Bradycardia	A0712	Pacing Problem	19/03/2021	19/03/2021
									A0908	Incorrect, Inadequate or Imprecise Result or Readings	19/03/2021	19/03/2021
							E2403	No Clinical Signs, Symptoms or Conditions	A0712	Pacing Problem	19/03/2021	19/03/2021
									A0908	Incorrect, Inadequate or Imprecise Result or Readings	19/03/2021	19/03/2021
							F08	Hospitalization or Prolonged Hospitalization	A0712	Pacing Problem	19/03/2021	19/03/2021
									A0908	Incorrect, Inadequate or Imprecise Result or Readings	19/03/2021	19/03/2021
							F12	Serious Injury/ Illness/ Impairment	A0712	Pacing Problem	19/03/2021	19/03/2021
									A0908	Incorrect, Inadequate or Imprecise Result or Readings	19/03/2021	19/03/2021
							F1905	Device Revision or Replacement	A0712	Pacing Problem	19/03/2021	19/03/2021
									A0908	Incorrect, Inadequate or Imprecise Result or Readings	19/03/2021	19/03/2021
					CT321	Long-term surgical invasive	E060104	Bradycardia	A0712	Pacing Problem	19/03/2021	19/03/2021
									A0908	Incorrect, Inadequate or Imprecise Result or Readings	19/03/2021	19/03/2021
							E2403	No Clinical Signs, Symptoms or Conditions	A0712	Pacing Problem	19/03/2021	19/03/2021
									A0908	Incorrect, Inadequate or Imprecise Result or Readings	19/03/2021	19/03/2021
							F08	Hospitalization or Prolonged Hospitalization	A0712	Pacing Problem	19/03/2021	19/03/2021
									A0908	Incorrect, Inadequate or Imprecise Result or Readings	19/03/2021	19/03/2021
							F12	Serious Injury/ Illness/ Impairment	A0712	Pacing Problem	19/03/2021	19/03/2021
									A0908	Incorrect, Inadequate or Imprecise Result or Readings	19/03/2021	19/03/2021
							F1905	Device Revision or Replacement	A0712	Pacing Problem	19/03/2021	19/03/2021
									A0908	Incorrect, Inadequate or Imprecise Result or Readings	19/03/2021	19/03/2021
					CT449	Generators and associated devices	E060104	Bradycardia	A0712	Pacing Problem	19/03/2021	19/03/2021
									A0908	Incorrect, Inadequate or Imprecise Result or Readings	19/03/2021	19/03/2021
							E2403	No Clinical Signs, Symptoms or Conditions	A0712	Pacing Problem	19/03/2021	19/03/2021
									A0908	Incorrect, Inadequate or Imprecise Result or Readings	19/03/2021	19/03/2021
							F08	Hospitalization or Prolonged Hospitalization	A0712	Pacing Problem	19/03/2021	19/03/2021
									A0908	Incorrect, Inadequate or Imprecise Result or Readings	19/03/2021	19/03/2021
							F12	Serious Injury/ Illness/ Impairment	A0712	Pacing Problem	19/03/2021	19/03/2021
									A0908	Incorrect, Inadequate or Imprecise Result or Readings	19/03/2021	19/03/2021
							F1905	Device Revision or Replacement	A0712	Pacing Problem	19/03/2021	19/03/2021

									A0908	Incorrect, Inadequate or Imprecise Result or Readings	19/03/2021	19/03/2021
					CT981	Single use	E060104	Bradycardia	A0712	Pacing Problem	19/03/2021	19/03/2021
									A0908	Incorrect, Inadequate or Imprecise Result or Readings	19/03/2021	19/03/2021
							E2403	No Clinical Signs, Symptoms or Conditions	A0712	Pacing Problem	19/03/2021	19/03/2021
									A0908	Incorrect, Inadequate or Imprecise Result or Readings	19/03/2021	19/03/2021
							F08	Hospitalization or Prolonged Hospitalization	A0712	Pacing Problem	19/03/2021	19/03/2021
									A0908	Incorrect, Inadequate or Imprecise Result or Readings	19/03/2021	19/03/2021
							F12	Serious Injury/ Illness/ Impairment	A0712	Pacing Problem	19/03/2021	19/03/2021
									A0908	Incorrect, Inadequate or Imprecise Result or Readings	19/03/2021	19/03/2021
							F1905	Device Revision or Replacement	A0712	Pacing Problem	19/03/2021	19/03/2021
									A0908	Incorrect, Inadequate or Imprecise Result or Readings	19/03/2021	19/03/2021
					CT983	Surgical invasive	E060104	Bradycardia	A0712	Pacing Problem	19/03/2021	19/03/2021
									A0908	Incorrect, Inadequate or Imprecise Result or Readings	19/03/2021	19/03/2021
							E2403	No Clinical Signs, Symptoms or Conditions	A0712	Pacing Problem	19/03/2021	19/03/2021
									A0908	Incorrect, Inadequate or Imprecise Result or Readings	19/03/2021	19/03/2021
							F08	Hospitalization or Prolonged Hospitalization	A0712	Pacing Problem	19/03/2021	19/03/2021
									A0908	Incorrect, Inadequate or Imprecise Result or Readings	19/03/2021	19/03/2021
							F12	Serious Injury/ Illness/ Impairment	A0712	Pacing Problem	19/03/2021	19/03/2021
									A0908	Incorrect, Inadequate or Imprecise Result or Readings	19/03/2021	19/03/2021
							F1905	Device Revision or Replacement	A0712	Pacing Problem	19/03/2021	19/03/2021
									A0908	Incorrect, Inadequate or Imprecise Result or Readings	19/03/2021	19/03/2021
10861	2021	All other reportable incidents	47265	Dual-chamber implantable pacemaker, rate-responsive	CT141	Electrodes/Leads and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	24/01/2022	24/01/2022
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	24/01/2022	24/01/2022
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	24/01/2022	24/01/2022
					CT145	Cardiovascular devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	24/01/2022	24/01/2022
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	24/01/2022	24/01/2022
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	24/01/2022	24/01/2022
					CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	24/01/2022	24/01/2022
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	24/01/2022	24/01/2022
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	24/01/2022	24/01/2022
					CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	24/01/2022	24/01/2022
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	24/01/2022	24/01/2022
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	24/01/2022	24/01/2022
					CT233	Surgical	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	24/01/2022	24/01/2022
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	24/01/2022	24/01/2022
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	24/01/2022	24/01/2022
					CT2406	Implantable	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	24/01/2022	24/01/2022
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	24/01/2022	24/01/2022
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	24/01/2022	24/01/2022
					CT2407	Active implantable	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	24/01/2022	24/01/2022
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	24/01/2022	24/01/2022
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	24/01/2022	24/01/2022
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	24/01/2022	24/01/2022
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	24/01/2022	24/01/2022
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	24/01/2022	24/01/2022
					CT2727	Active implant devices/ancillaries	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	24/01/2022	24/01/2022
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	24/01/2022	24/01/2022
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	24/01/2022	24/01/2022
					CT2750	Implantable cardiac electrical stimulators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	24/01/2022	24/01/2022
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	24/01/2022	24/01/2022
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	24/01/2022	24/01/2022
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	24/01/2022	24/01/2022
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	24/01/2022	24/01/2022
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	24/01/2022	24/01/2022
					CT2985	Active	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	24/01/2022	24/01/2022
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	24/01/2022	24/01/2022
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	24/01/2022	24/01/2022
					CT3001	Implantable metal	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	24/01/2022	24/01/2022
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	24/01/2022	24/01/2022
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	24/01/2022	24/01/2022
					CT3002	Implantable synthetic polymer	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	24/01/2022	24/01/2022
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	24/01/2022	24/01/2022
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	24/01/2022	24/01/2022
					CT3006	Implantable inorganic material	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	24/01/2022	24/01/2022
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	24/01/2022	24/01/2022
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	24/01/2022	24/01/2022
					CT314	Electrically powered	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	24/01/2022	24/01/2022
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	24/01/2022	24/01/2022
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	24/01/2022	24/01/2022
					CT321	Long-term surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	24/01/2022	24/01/2022
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	24/01/2022	24/01/2022
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	24/01/2022	24/01/2022
					CT449	Generators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	24/01/2022	24/01/2022
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	24/01/2022	24/01/2022
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	24/01/2022	24/01/2022
					CT981	Single use	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	24/01/2022	24/01/2022
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	24/01/2022	24/01/2022
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	24/01/2022	24/01/2022
					CT983	Surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	24/01/2022	24/01/2022

							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	24/01/2022	24/01/2022
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	24/01/2022	24/01/2022
10862	2021	All other reportable incidents	47265	Dual-chamber implantable pacemaker, rate-responsive	CT141	Electrodes/Leads and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A0203	Defective Device		
									A070504	Premature Discharge of Battery		
							F08	Hospitalization or Prolonged Hospitalization	A0203	Defective Device		
									A070504	Premature Discharge of Battery		
							F1905	Device Revision or Replacement	A0203	Defective Device		
									A070504	Premature Discharge of Battery		
					CT145	Cardiovascular devices	E2403	No Clinical Signs, Symptoms or Conditions	A0203	Defective Device		
									A070504	Premature Discharge of Battery		
							F08	Hospitalization or Prolonged Hospitalization	A0203	Defective Device		
									A070504	Premature Discharge of Battery		
							F1905	Device Revision or Replacement	A0203	Defective Device		
									A070504	Premature Discharge of Battery		
					CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A0203	Defective Device		
									A070504	Premature Discharge of Battery		
							F08	Hospitalization or Prolonged Hospitalization	A0203	Defective Device		
									A070504	Premature Discharge of Battery		
							F1905	Device Revision or Replacement	A0203	Defective Device		
									A070504	Premature Discharge of Battery		
					CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A0203	Defective Device		
									A070504	Premature Discharge of Battery		
							F08	Hospitalization or Prolonged Hospitalization	A0203	Defective Device		
									A070504	Premature Discharge of Battery		
							F1905	Device Revision or Replacement	A0203	Defective Device		
									A070504	Premature Discharge of Battery		
					CT233	Surgical	E2403	No Clinical Signs, Symptoms or Conditions	A0203	Defective Device		
									A070504	Premature Discharge of Battery		
							F08	Hospitalization or Prolonged Hospitalization	A0203	Defective Device		
									A070504	Premature Discharge of Battery		
							F1905	Device Revision or Replacement	A0203	Defective Device		
									A070504	Premature Discharge of Battery		
					CT2406	Implantable	E2403	No Clinical Signs, Symptoms or Conditions	A0203	Defective Device		
									A070504	Premature Discharge of Battery		
							F08	Hospitalization or Prolonged Hospitalization	A0203	Defective Device		
									A070504	Premature Discharge of Battery		
							F1905	Device Revision or Replacement	A0203	Defective Device		
									A070504	Premature Discharge of Battery		
					CT2407	Active implantable	E2403	No Clinical Signs, Symptoms or Conditions	A0203	Defective Device		
									A070504	Premature Discharge of Battery		
							F08	Hospitalization or Prolonged Hospitalization	A0203	Defective Device		
									A070504	Premature Discharge of Battery		
							F1905	Device Revision or Replacement	A0203	Defective Device		
									A070504	Premature Discharge of Battery		
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A0203	Defective Device		
									A070504	Premature Discharge of Battery		
							F08	Hospitalization or Prolonged Hospitalization	A0203	Defective Device		
									A070504	Premature Discharge of Battery		
							F1905	Device Revision or Replacement	A0203	Defective Device		
									A070504	Premature Discharge of Battery		
					CT2727	Active implant devices/ancillaries	E2403	No Clinical Signs, Symptoms or Conditions	A0203	Defective Device		
									A070504	Premature Discharge of Battery		
							F08	Hospitalization or Prolonged Hospitalization	A0203	Defective Device		
									A070504	Premature Discharge of Battery		
							F1905	Device Revision or Replacement	A0203	Defective Device		
									A070504	Premature Discharge of Battery		
					CT2750	Implantable cardiac electrical stimulators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A0203	Defective Device		
									A070504	Premature Discharge of Battery		
							F08	Hospitalization or Prolonged Hospitalization	A0203	Defective Device		
									A070504	Premature Discharge of Battery		
							F1905	Device Revision or Replacement	A0203	Defective Device		
									A070504	Premature Discharge of Battery		
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A0203	Defective Device		
									A070504	Premature Discharge of Battery		
							F08	Hospitalization or Prolonged Hospitalization	A0203	Defective Device		
									A070504	Premature Discharge of Battery		
							F1905	Device Revision or Replacement	A0203	Defective Device		
									A070504	Premature Discharge of Battery		
					CT2985	Active	E2403	No Clinical Signs, Symptoms or Conditions	A0203	Defective Device		
									A070504	Premature Discharge of Battery		
							F08	Hospitalization or Prolonged Hospitalization	A0203	Defective Device		
									A070504	Premature Discharge of Battery		
							F1905	Device Revision or Replacement	A0203	Defective Device		
									A070504	Premature Discharge of Battery		
					CT3001	Implantable metal	E2403	No Clinical Signs, Symptoms or Conditions	A0203	Defective Device		
									A070504	Premature Discharge of Battery		
							F08	Hospitalization or Prolonged Hospitalization	A0203	Defective Device		
									A070504	Premature Discharge of Battery		
							F1905	Device Revision or Replacement	A0203	Defective Device		
									A070504	Premature Discharge of Battery		
					CT3002	Implantable synthetic polymer	E2403	No Clinical Signs, Symptoms or Conditions	A0203	Defective Device		
									A070504	Premature Discharge of Battery		
							F08	Hospitalization or Prolonged Hospitalization	A0203	Defective Device		
									A070504	Premature Discharge of Battery		
							F1905	Device Revision or Replacement	A0203	Defective Device		
									A070504	Premature Discharge of Battery		
					CT3006	Implantable inorganic material	E2403	No Clinical Signs, Symptoms or Conditions	A0203	Defective Device		
									A070504	Premature Discharge of Battery		

						F08	Hospitalization or Prolonged Hospitalization	A0203	Defective Device		
								A070504	Premature Discharge of Battery		
						F1905	Device Revision or Replacement	A0203	Defective Device		
								A070504	Premature Discharge of Battery		
					CT314	Electrically powered	E2403	No Clinical Signs, Symptoms or Conditions	A0203	Defective Device		
									A070504	Premature Discharge of Battery		
							F08	Hospitalization or Prolonged Hospitalization	A0203	Defective Device		
									A070504	Premature Discharge of Battery		
							F1905	Device Revision or Replacement	A0203	Defective Device		
									A070504	Premature Discharge of Battery		
					CT321	Long-term surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A0203	Defective Device		
									A070504	Premature Discharge of Battery		
							F08	Hospitalization or Prolonged Hospitalization	A0203	Defective Device		
									A070504	Premature Discharge of Battery		
							F1905	Device Revision or Replacement	A0203	Defective Device		
									A070504	Premature Discharge of Battery		
					CT449	Generators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A0203	Defective Device		
									A070504	Premature Discharge of Battery		
							F08	Hospitalization or Prolonged Hospitalization	A0203	Defective Device		
									A070504	Premature Discharge of Battery		
							F1905	Device Revision or Replacement	A0203	Defective Device		
									A070504	Premature Discharge of Battery		
					CT981	Single use	E2403	No Clinical Signs, Symptoms or Conditions	A0203	Defective Device		
									A070504	Premature Discharge of Battery		
							F08	Hospitalization or Prolonged Hospitalization	A0203	Defective Device		
									A070504	Premature Discharge of Battery		
							F1905	Device Revision or Replacement	A0203	Defective Device		
									A070504	Premature Discharge of Battery		
					CT983	Surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A0203	Defective Device		
									A070504	Premature Discharge of Battery		
							F08	Hospitalization or Prolonged Hospitalization	A0203	Defective Device		
									A070504	Premature Discharge of Battery		
							F1905	Device Revision or Replacement	A0203	Defective Device		
									A070504	Premature Discharge of Battery		
10863	2021	All other reportable incidents	47267	Single-chamber implantable pacemaker, rate-responsive	CT141	Electrodes/Leads and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
							F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
					CT145	Cardiovascular devices	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
							F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
					CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
							F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
					CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
							F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
					CT233	Surgical	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
							F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
					CT2406	Implantable	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
							F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
					CT2407	Active implantable	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
							F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
							F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
					CT2727	Active implant devices/ancillaries	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
							F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		

					CT2750	Implantable cardiac electrical stimulators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
							F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
							F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
					CT2985	Active	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
							F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
					CT3001	Implantable metal	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
							F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
					CT3002	Implantable synthetic polymer	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
							F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
					CT3006	Implantable inorganic material	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
							F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
					CT314	Electrically powered	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
							F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
					CT321	Long-term surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
							F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
					CT449	Generators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
							F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
					CT981	Single use	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
							F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
					CT983	Surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
							F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
10864	2021	All other reportable incidents	36241	Pacemaker/defibrillator lead	CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A0203	Defective Device	24/03/2021	24/03/2021
							F1908	Prolonged Surgery	A0203	Defective Device	24/03/2021	24/03/2021
							F26	No Health Consequences or Impact	A0203	Defective Device	24/03/2021	24/03/2021
10865	2021	All other reportable incidents	47263	Cardiac resynchronization therapy implantable pacemaker	CT1684	Cardiac pacemakers and associated devices	E0112	Dizziness	A071202	Inaccurate Synchronization	26/03/2021	26/03/2021
							E0601	Arrhythmia	A071202	Inaccurate Synchronization	26/03/2021	26/03/2021
							E2401	Insufficient Information	A071202	Inaccurate Synchronization	26/03/2021	26/03/2021
							F08	Hospitalization or Prolonged Hospitalization	A071202	Inaccurate Synchronization	26/03/2021	26/03/2021
							F1905	Device Revision or Replacement	A071202	Inaccurate Synchronization	26/03/2021	26/03/2021
					CT1685	Cardiac pacemakers	E0112	Dizziness	A071202	Inaccurate Synchronization	26/03/2021	26/03/2021
							E0601	Arrhythmia	A071202	Inaccurate Synchronization	26/03/2021	26/03/2021
							E2401	Insufficient Information	A071202	Inaccurate Synchronization	26/03/2021	26/03/2021
							F08	Hospitalization or Prolonged Hospitalization	A071202	Inaccurate Synchronization	26/03/2021	26/03/2021
							F1905	Device Revision or Replacement	A071202	Inaccurate Synchronization	26/03/2021	26/03/2021
					CT257	Implantable cardiac pacemakers	E0112	Dizziness	A071202	Inaccurate Synchronization	26/03/2021	26/03/2021
							E0601	Arrhythmia	A071202	Inaccurate Synchronization	26/03/2021	26/03/2021
							E2401	Insufficient Information	A071202	Inaccurate Synchronization	26/03/2021	26/03/2021
							F08	Hospitalization or Prolonged Hospitalization	A071202	Inaccurate Synchronization	26/03/2021	26/03/2021
							F1905	Device Revision or Replacement	A071202	Inaccurate Synchronization	26/03/2021	26/03/2021
					CT2866	Implantable cardiac electrical stimulators	E0112	Dizziness	A071202	Inaccurate Synchronization	26/03/2021	26/03/2021
							E0601	Arrhythmia	A071202	Inaccurate Synchronization	26/03/2021	26/03/2021
							E2401	Insufficient Information	A071202	Inaccurate Synchronization	26/03/2021	26/03/2021
							F08	Hospitalization or Prolonged Hospitalization	A071202	Inaccurate Synchronization	26/03/2021	26/03/2021

							F1905	Device Revision or Replacement	A071202	Inaccurate Synchronization	26/03/2021	26/03/2021
10866	2021	All other reportable incidents	47265	Dual-chamber implantable pacemaker, rate-responsive	CT141	Electrodes/Leads and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT145	Cardiovascular devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT233	Surgical	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT2406	Implantable	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT2407	Active implantable	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT2727	Active implant devices/ancillaries	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT2750	Implantable cardiac electrical stimulators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT2985	Active	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT3001	Implantable metal	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT3002	Implantable synthetic polymer	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT3006	Implantable inorganic material	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT314	Electrically powered	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT321	Long-term surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT449	Generators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT981	Single use	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT983	Surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
10867	2021	All other reportable incidents	47265	Dual-chamber implantable pacemaker, rate-responsive	CT141	Electrodes/Leads and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT145	Cardiovascular devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT233	Surgical	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT2406	Implantable	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT2407	Active implantable	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT2727	Active implant devices/ancillaries	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT2750	Implantable cardiac electrical stimulators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT2985	Active	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT3001	Implantable metal	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT3002	Implantable synthetic polymer	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT3006	Implantable inorganic material	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT314	Electrically powered	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT321	Long-term surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT449	Generators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT981	Single use	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT983	Surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
10868	2021	All other reportable incidents	17882	Physiologic-monitoring defibrillation system	CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070803	Failure to Power Up		
							F27	Problem Identified During Non-Clinical Procedure	A070803	Failure to Power Up		
10869	2021	All other reportable incidents	47265	Dual-chamber implantable pacemaker, rate-responsive	CT141	Electrodes/Leads and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070909	Oversensing	01/01/2019	31/12/2019
									A0712	Pacing Problem	01/01/2019	31/12/2019
							F1901	Additional Surgery	A070909	Oversensing	01/01/2019	31/12/2019
									A0712	Pacing Problem	01/01/2019	31/12/2019
							F1905	Device Revision or Replacement	A070909	Oversensing	01/01/2019	31/12/2019

								A0712	Pacing Problem	01/01/2019	31/12/2019
					CT145	Cardiovascular devices	E2403	No Clinical Signs, Symptoms or Conditions	A070909	Oversensing	01/01/2019	31/12/2019
									A0712	Pacing Problem	01/01/2019	31/12/2019
							F1901	Additional Surgery	A070909	Oversensing	01/01/2019	31/12/2019
									A0712	Pacing Problem	01/01/2019	31/12/2019
							F1905	Device Revision or Replacement	A070909	Oversensing	01/01/2019	31/12/2019
									A0712	Pacing Problem	01/01/2019	31/12/2019
					CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070909	Oversensing	01/01/2019	31/12/2019
									A0712	Pacing Problem	01/01/2019	31/12/2019
							F1901	Additional Surgery	A070909	Oversensing	01/01/2019	31/12/2019
									A0712	Pacing Problem	01/01/2019	31/12/2019
							F1905	Device Revision or Replacement	A070909	Oversensing	01/01/2019	31/12/2019
									A0712	Pacing Problem	01/01/2019	31/12/2019
					CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A070909	Oversensing	01/01/2019	31/12/2019
									A0712	Pacing Problem	01/01/2019	31/12/2019
							F1901	Additional Surgery	A070909	Oversensing	01/01/2019	31/12/2019
									A0712	Pacing Problem	01/01/2019	31/12/2019
							F1905	Device Revision or Replacement	A070909	Oversensing	01/01/2019	31/12/2019
									A0712	Pacing Problem	01/01/2019	31/12/2019
					CT233	Surgical	E2403	No Clinical Signs, Symptoms or Conditions	A070909	Oversensing	01/01/2019	31/12/2019
									A0712	Pacing Problem	01/01/2019	31/12/2019
							F1901	Additional Surgery	A070909	Oversensing	01/01/2019	31/12/2019
									A0712	Pacing Problem	01/01/2019	31/12/2019
							F1905	Device Revision or Replacement	A070909	Oversensing	01/01/2019	31/12/2019
									A0712	Pacing Problem	01/01/2019	31/12/2019
					CT2406	Implantable	E2403	No Clinical Signs, Symptoms or Conditions	A070909	Oversensing	01/01/2019	31/12/2019
									A0712	Pacing Problem	01/01/2019	31/12/2019
							F1901	Additional Surgery	A070909	Oversensing	01/01/2019	31/12/2019
									A0712	Pacing Problem	01/01/2019	31/12/2019
							F1905	Device Revision or Replacement	A070909	Oversensing	01/01/2019	31/12/2019
									A0712	Pacing Problem	01/01/2019	31/12/2019
					CT2407	Active implantable	E2403	No Clinical Signs, Symptoms or Conditions	A070909	Oversensing	01/01/2019	31/12/2019
									A0712	Pacing Problem	01/01/2019	31/12/2019
							F1901	Additional Surgery	A070909	Oversensing	01/01/2019	31/12/2019
									A0712	Pacing Problem	01/01/2019	31/12/2019
							F1905	Device Revision or Replacement	A070909	Oversensing	01/01/2019	31/12/2019
									A0712	Pacing Problem	01/01/2019	31/12/2019
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A070909	Oversensing	01/01/2019	31/12/2019
									A0712	Pacing Problem	01/01/2019	31/12/2019
							F1901	Additional Surgery	A070909	Oversensing	01/01/2019	31/12/2019
									A0712	Pacing Problem	01/01/2019	31/12/2019
							F1905	Device Revision or Replacement	A070909	Oversensing	01/01/2019	31/12/2019
									A0712	Pacing Problem	01/01/2019	31/12/2019
					CT2727	Active implant devices/ancillaries	E2403	No Clinical Signs, Symptoms or Conditions	A070909	Oversensing	01/01/2019	31/12/2019
									A0712	Pacing Problem	01/01/2019	31/12/2019
							F1901	Additional Surgery	A070909	Oversensing	01/01/2019	31/12/2019
									A0712	Pacing Problem	01/01/2019	31/12/2019
							F1905	Device Revision or Replacement	A070909	Oversensing	01/01/2019	31/12/2019
									A0712	Pacing Problem	01/01/2019	31/12/2019
					CT2750	Implantable cardiac electrical stimulators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070909	Oversensing	01/01/2019	31/12/2019
									A0712	Pacing Problem	01/01/2019	31/12/2019
							F1901	Additional Surgery	A070909	Oversensing	01/01/2019	31/12/2019
									A0712	Pacing Problem	01/01/2019	31/12/2019
							F1905	Device Revision or Replacement	A070909	Oversensing	01/01/2019	31/12/2019
									A0712	Pacing Problem	01/01/2019	31/12/2019
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A070909	Oversensing	01/01/2019	31/12/2019
									A0712	Pacing Problem	01/01/2019	31/12/2019
							F1901	Additional Surgery	A070909	Oversensing	01/01/2019	31/12/2019
									A0712	Pacing Problem	01/01/2019	31/12/2019
							F1905	Device Revision or Replacement	A070909	Oversensing	01/01/2019	31/12/2019
									A0712	Pacing Problem	01/01/2019	31/12/2019
					CT2985	Active	E2403	No Clinical Signs, Symptoms or Conditions	A070909	Oversensing	01/01/2019	31/12/2019
									A0712	Pacing Problem	01/01/2019	31/12/2019
							F1901	Additional Surgery	A070909	Oversensing	01/01/2019	31/12/2019
									A0712	Pacing Problem	01/01/2019	31/12/2019
							F1905	Device Revision or Replacement	A070909	Oversensing	01/01/2019	31/12/2019
									A0712	Pacing Problem	01/01/2019	31/12/2019
					CT3001	Implantable metal	E2403	No Clinical Signs, Symptoms or Conditions	A070909	Oversensing	01/01/2019	31/12/2019
									A0712	Pacing Problem	01/01/2019	31/12/2019
							F1901	Additional Surgery	A070909	Oversensing	01/01/2019	31/12/2019
									A0712	Pacing Problem	01/01/2019	31/12/2019
							F1905	Device Revision or Replacement	A070909	Oversensing	01/01/2019	31/12/2019
									A0712	Pacing Problem	01/01/2019	31/12/2019
					CT3002	Implantable synthetic polymer	E2403	No Clinical Signs, Symptoms or Conditions	A070909	Oversensing	01/01/2019	31/12/2019
									A0712	Pacing Problem	01/01/2019	31/12/2019
							F1901	Additional Surgery	A070909	Oversensing	01/01/2019	31/12/2019
									A0712	Pacing Problem	01/01/2019	31/12/2019
							F1905	Device Revision or Replacement	A070909	Oversensing	01/01/2019	31/12/2019
									A0712	Pacing Problem	01/01/2019	31/12/2019
					CT3006	Implantable inorganic material	E2403	No Clinical Signs, Symptoms or Conditions	A070909	Oversensing	01/01/2019	31/12/2019
									A0712	Pacing Problem	01/01/2019	31/12/2019
							F1901	Additional Surgery	A070909	Oversensing	01/01/2019	31/12/2019
									A0712	Pacing Problem	01/01/2019	31/12/2019
							F1905	Device Revision or Replacement	A070909	Oversensing	01/01/2019	31/12/2019
									A0712	Pacing Problem	01/01/2019	31/12/2019
					CT314	Electrically powered	E2403	No Clinical Signs, Symptoms or Conditions	A070909	Oversensing	01/01/2019	31/12/2019
									A0712	Pacing Problem	01/01/2019	31/12/2019
							F1901	Additional Surgery	A070909	Oversensing	01/01/2019	31/12/2019

								A0712	Pacing Problem	01/01/2019	31/12/2019
						F1905	Device Revision or Replacement	A070909	Oversensing	01/01/2019	31/12/2019
								A0712	Pacing Problem	01/01/2019	31/12/2019
					CT321	Long-term surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A070909	Oversensing	01/01/2019	31/12/2019
									A0712	Pacing Problem	01/01/2019	31/12/2019
							F1901	Additional Surgery	A070909	Oversensing	01/01/2019	31/12/2019
									A0712	Pacing Problem	01/01/2019	31/12/2019
							F1905	Device Revision or Replacement	A070909	Oversensing	01/01/2019	31/12/2019
									A0712	Pacing Problem	01/01/2019	31/12/2019
					CT449	Generators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070909	Oversensing	01/01/2019	31/12/2019
									A0712	Pacing Problem	01/01/2019	31/12/2019
							F1901	Additional Surgery	A070909	Oversensing	01/01/2019	31/12/2019
									A0712	Pacing Problem	01/01/2019	31/12/2019
							F1905	Device Revision or Replacement	A070909	Oversensing	01/01/2019	31/12/2019
									A0712	Pacing Problem	01/01/2019	31/12/2019
					CT981	Single use	E2403	No Clinical Signs, Symptoms or Conditions	A070909	Oversensing	01/01/2019	31/12/2019
									A0712	Pacing Problem	01/01/2019	31/12/2019
							F1901	Additional Surgery	A070909	Oversensing	01/01/2019	31/12/2019
									A0712	Pacing Problem	01/01/2019	31/12/2019
							F1905	Device Revision or Replacement	A070909	Oversensing	01/01/2019	31/12/2019
									A0712	Pacing Problem	01/01/2019	31/12/2019
					CT983	Surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A070909	Oversensing	01/01/2019	31/12/2019
									A0712	Pacing Problem	01/01/2019	31/12/2019
							F1901	Additional Surgery	A070909	Oversensing	01/01/2019	31/12/2019
									A0712	Pacing Problem	01/01/2019	31/12/2019
							F1905	Device Revision or Replacement	A070909	Oversensing	01/01/2019	31/12/2019
									A0712	Pacing Problem	01/01/2019	31/12/2019
10870	2021	All other reportable incidents	47265	Dual-chamber implantable pacemaker, rate-responsive	CT141	Electrodes/Leads and associated devices	E011903	Syncope/Fainting	A070909	Oversensing	01/02/2019	01/02/2019
									A0712	Pacing Problem	01/02/2019	01/02/2019
							F1901	Additional Surgery	A070909	Oversensing	01/02/2019	01/02/2019
									A0712	Pacing Problem	01/02/2019	01/02/2019
							F1905	Device Revision or Replacement	A070909	Oversensing	01/02/2019	01/02/2019
									A0712	Pacing Problem	01/02/2019	01/02/2019
							F21	Unexpected Deterioration	A070909	Oversensing	01/02/2019	01/02/2019
									A0712	Pacing Problem	01/02/2019	01/02/2019
					CT145	Cardiovascular devices	E011903	Syncope/Fainting	A070909	Oversensing	01/02/2019	01/02/2019
									A0712	Pacing Problem	01/02/2019	01/02/2019
							F1901	Additional Surgery	A070909	Oversensing	01/02/2019	01/02/2019
									A0712	Pacing Problem	01/02/2019	01/02/2019
							F1905	Device Revision or Replacement	A070909	Oversensing	01/02/2019	01/02/2019
									A0712	Pacing Problem	01/02/2019	01/02/2019
							F21	Unexpected Deterioration	A070909	Oversensing	01/02/2019	01/02/2019
									A0712	Pacing Problem	01/02/2019	01/02/2019
					CT1684	Cardiac pacemakers and associated devices	E011903	Syncope/Fainting	A070909	Oversensing	01/02/2019	01/02/2019
									A0712	Pacing Problem	01/02/2019	01/02/2019
							F1901	Additional Surgery	A070909	Oversensing	01/02/2019	01/02/2019
									A0712	Pacing Problem	01/02/2019	01/02/2019
							F1905	Device Revision or Replacement	A070909	Oversensing	01/02/2019	01/02/2019
									A0712	Pacing Problem	01/02/2019	01/02/2019
							F21	Unexpected Deterioration	A070909	Oversensing	01/02/2019	01/02/2019
									A0712	Pacing Problem	01/02/2019	01/02/2019
					CT1685	Cardiac pacemakers	E011903	Syncope/Fainting	A070909	Oversensing	01/02/2019	01/02/2019
									A0712	Pacing Problem	01/02/2019	01/02/2019
							F1901	Additional Surgery	A070909	Oversensing	01/02/2019	01/02/2019
									A0712	Pacing Problem	01/02/2019	01/02/2019
							F1905	Device Revision or Replacement	A070909	Oversensing	01/02/2019	01/02/2019
									A0712	Pacing Problem	01/02/2019	01/02/2019
							F21	Unexpected Deterioration	A070909	Oversensing	01/02/2019	01/02/2019
									A0712	Pacing Problem	01/02/2019	01/02/2019
					CT233	Surgical	E011903	Syncope/Fainting	A070909	Oversensing	01/02/2019	01/02/2019
									A0712	Pacing Problem	01/02/2019	01/02/2019
							F1901	Additional Surgery	A070909	Oversensing	01/02/2019	01/02/2019
									A0712	Pacing Problem	01/02/2019	01/02/2019
							F1905	Device Revision or Replacement	A070909	Oversensing	01/02/2019	01/02/2019
									A0712	Pacing Problem	01/02/2019	01/02/2019
							F21	Unexpected Deterioration	A070909	Oversensing	01/02/2019	01/02/2019
									A0712	Pacing Problem	01/02/2019	01/02/2019
					CT2406	Implantable	E011903	Syncope/Fainting	A070909	Oversensing	01/02/2019	01/02/2019
									A0712	Pacing Problem	01/02/2019	01/02/2019
							F1901	Additional Surgery	A070909	Oversensing	01/02/2019	01/02/2019
									A0712	Pacing Problem	01/02/2019	01/02/2019
							F1905	Device Revision or Replacement	A070909	Oversensing	01/02/2019	01/02/2019
									A0712	Pacing Problem	01/02/2019	01/02/2019
							F21	Unexpected Deterioration	A070909	Oversensing	01/02/2019	01/02/2019
									A0712	Pacing Problem	01/02/2019	01/02/2019
					CT2407	Active implantable	E011903	Syncope/Fainting	A070909	Oversensing	01/02/2019	01/02/2019
									A0712	Pacing Problem	01/02/2019	01/02/2019
							F1901	Additional Surgery	A070909	Oversensing	01/02/2019	01/02/2019
									A0712	Pacing Problem	01/02/2019	01/02/2019
							F1905	Device Revision or Replacement	A070909	Oversensing	01/02/2019	01/02/2019
									A0712	Pacing Problem	01/02/2019	01/02/2019
							F21	Unexpected Deterioration	A070909	Oversensing	01/02/2019	01/02/2019
									A0712	Pacing Problem	01/02/2019	01/02/2019
					CT257	Implantable cardiac pacemakers	E011903	Syncope/Fainting	A070909	Oversensing	01/02/2019	01/02/2019
									A0712	Pacing Problem	01/02/2019	01/02/2019
							F1901	Additional Surgery	A070909	Oversensing	01/02/2019	01/02/2019
									A0712	Pacing Problem	01/02/2019	01/02/2019
							F1905	Device Revision or Replacement	A070909	Oversensing	01/02/2019	01/02/2019

									A0712	Pacing Problem	01/02/2019	01/02/2019
							F21	Unexpected Deterioration	A070909	Oversensing	01/02/2019	01/02/2019
									A0712	Pacing Problem	01/02/2019	01/02/2019
					CT2727	Active implant devices/ancillaries	E011903	Syncope/Fainting	A070909	Oversensing	01/02/2019	01/02/2019
									A0712	Pacing Problem	01/02/2019	01/02/2019
							F1901	Additional Surgery	A070909	Oversensing	01/02/2019	01/02/2019
									A0712	Pacing Problem	01/02/2019	01/02/2019
							F1905	Device Revision or Replacement	A070909	Oversensing	01/02/2019	01/02/2019
									A0712	Pacing Problem	01/02/2019	01/02/2019
							F21	Unexpected Deterioration	A070909	Oversensing	01/02/2019	01/02/2019
									A0712	Pacing Problem	01/02/2019	01/02/2019
					CT2750	Implantable cardiac electrical stimulators and associated devices	E011903	Syncope/Fainting	A070909	Oversensing	01/02/2019	01/02/2019
									A0712	Pacing Problem	01/02/2019	01/02/2019
							F1901	Additional Surgery	A070909	Oversensing	01/02/2019	01/02/2019
									A0712	Pacing Problem	01/02/2019	01/02/2019
							F1905	Device Revision or Replacement	A070909	Oversensing	01/02/2019	01/02/2019
									A0712	Pacing Problem	01/02/2019	01/02/2019
							F21	Unexpected Deterioration	A070909	Oversensing	01/02/2019	01/02/2019
									A0712	Pacing Problem	01/02/2019	01/02/2019
					CT2866	Implantable cardiac electrical stimulators	E011903	Syncope/Fainting	A070909	Oversensing	01/02/2019	01/02/2019
									A0712	Pacing Problem	01/02/2019	01/02/2019
							F1901	Additional Surgery	A070909	Oversensing	01/02/2019	01/02/2019
									A0712	Pacing Problem	01/02/2019	01/02/2019
							F1905	Device Revision or Replacement	A070909	Oversensing	01/02/2019	01/02/2019
									A0712	Pacing Problem	01/02/2019	01/02/2019
							F21	Unexpected Deterioration	A070909	Oversensing	01/02/2019	01/02/2019
									A0712	Pacing Problem	01/02/2019	01/02/2019
					CT2985	Active	E011903	Syncope/Fainting	A070909	Oversensing	01/02/2019	01/02/2019
									A0712	Pacing Problem	01/02/2019	01/02/2019
							F1901	Additional Surgery	A070909	Oversensing	01/02/2019	01/02/2019
									A0712	Pacing Problem	01/02/2019	01/02/2019
							F1905	Device Revision or Replacement	A070909	Oversensing	01/02/2019	01/02/2019
									A0712	Pacing Problem	01/02/2019	01/02/2019
							F21	Unexpected Deterioration	A070909	Oversensing	01/02/2019	01/02/2019
									A0712	Pacing Problem	01/02/2019	01/02/2019
					CT3001	Implantable metal	E011903	Syncope/Fainting	A070909	Oversensing	01/02/2019	01/02/2019
									A0712	Pacing Problem	01/02/2019	01/02/2019
							F1901	Additional Surgery	A070909	Oversensing	01/02/2019	01/02/2019
									A0712	Pacing Problem	01/02/2019	01/02/2019
							F1905	Device Revision or Replacement	A070909	Oversensing	01/02/2019	01/02/2019
									A0712	Pacing Problem	01/02/2019	01/02/2019
							F21	Unexpected Deterioration	A070909	Oversensing	01/02/2019	01/02/2019
									A0712	Pacing Problem	01/02/2019	01/02/2019
					CT3002	Implantable synthetic polymer	E011903	Syncope/Fainting	A070909	Oversensing	01/02/2019	01/02/2019
									A0712	Pacing Problem	01/02/2019	01/02/2019
							F1901	Additional Surgery	A070909	Oversensing	01/02/2019	01/02/2019
									A0712	Pacing Problem	01/02/2019	01/02/2019
							F1905	Device Revision or Replacement	A070909	Oversensing	01/02/2019	01/02/2019
									A0712	Pacing Problem	01/02/2019	01/02/2019
							F21	Unexpected Deterioration	A070909	Oversensing	01/02/2019	01/02/2019
									A0712	Pacing Problem	01/02/2019	01/02/2019
					CT3006	Implantable inorganic material	E011903	Syncope/Fainting	A070909	Oversensing	01/02/2019	01/02/2019
									A0712	Pacing Problem	01/02/2019	01/02/2019
							F1901	Additional Surgery	A070909	Oversensing	01/02/2019	01/02/2019
									A0712	Pacing Problem	01/02/2019	01/02/2019
							F1905	Device Revision or Replacement	A070909	Oversensing	01/02/2019	01/02/2019
									A0712	Pacing Problem	01/02/2019	01/02/2019
							F21	Unexpected Deterioration	A070909	Oversensing	01/02/2019	01/02/2019
									A0712	Pacing Problem	01/02/2019	01/02/2019
					CT314	Electrically powered	E011903	Syncope/Fainting	A070909	Oversensing	01/02/2019	01/02/2019
									A0712	Pacing Problem	01/02/2019	01/02/2019
							F1901	Additional Surgery	A070909	Oversensing	01/02/2019	01/02/2019
									A0712	Pacing Problem	01/02/2019	01/02/2019
							F1905	Device Revision or Replacement	A070909	Oversensing	01/02/2019	01/02/2019
									A0712	Pacing Problem	01/02/2019	01/02/2019
							F21	Unexpected Deterioration	A070909	Oversensing	01/02/2019	01/02/2019
									A0712	Pacing Problem	01/02/2019	01/02/2019
					CT321	Long-term surgical invasive	E011903	Syncope/Fainting	A070909	Oversensing	01/02/2019	01/02/2019
									A0712	Pacing Problem	01/02/2019	01/02/2019
							F1901	Additional Surgery	A070909	Oversensing	01/02/2019	01/02/2019
									A0712	Pacing Problem	01/02/2019	01/02/2019
							F1905	Device Revision or Replacement	A070909	Oversensing	01/02/2019	01/02/2019
									A0712	Pacing Problem	01/02/2019	01/02/2019
							F21	Unexpected Deterioration	A070909	Oversensing	01/02/2019	01/02/2019
									A0712	Pacing Problem	01/02/2019	01/02/2019
					CT449	Generators and associated devices	E011903	Syncope/Fainting	A070909	Oversensing	01/02/2019	01/02/2019
									A0712	Pacing Problem	01/02/2019	01/02/2019
							F1901	Additional Surgery	A070909	Oversensing	01/02/2019	01/02/2019
									A0712	Pacing Problem	01/02/2019	01/02/2019
							F1905	Device Revision or Replacement	A070909	Oversensing	01/02/2019	01/02/2019
									A0712	Pacing Problem	01/02/2019	01/02/2019
							F21	Unexpected Deterioration	A070909	Oversensing	01/02/2019	01/02/2019
									A0712	Pacing Problem	01/02/2019	01/02/2019
					CT981	Single use	E011903	Syncope/Fainting	A070909	Oversensing	01/02/2019	01/02/2019
									A0712	Pacing Problem	01/02/2019	01/02/2019
							F1901	Additional Surgery	A070909	Oversensing	01/02/2019	01/02/2019
									A0712	Pacing Problem	01/02/2019	01/02/2019
							F1905	Device Revision or Replacement	A070909	Oversensing	01/02/2019	01/02/2019

									A0712	Pacing Problem	01/02/2019	01/02/2019
							F21	Unexpected Deterioration	A070909	Oversensing	01/02/2019	01/02/2019
									A0712	Pacing Problem	01/02/2019	01/02/2019
					CT983	Surgical invasive	E011903	Syncope/Fainting	A070909	Oversensing	01/02/2019	01/02/2019
									A0712	Pacing Problem	01/02/2019	01/02/2019
							F1901	Additional Surgery	A070909	Oversensing	01/02/2019	01/02/2019
									A0712	Pacing Problem	01/02/2019	01/02/2019
							F1905	Device Revision or Replacement	A070909	Oversensing	01/02/2019	01/02/2019
									A0712	Pacing Problem	01/02/2019	01/02/2019
							F21	Unexpected Deterioration	A070909	Oversensing	01/02/2019	01/02/2019
									A0712	Pacing Problem	01/02/2019	01/02/2019
10871	2021	All other reportable incidents	47270	Cardiac resynchronization therapy implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
10872	2021	All other reportable incidents	60789	Intracardiac pacemaker	CT130	Catheters and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	30/04/2021	30/04/2021
									A071102	Failure to Interrogate	30/04/2021	30/04/2021
							F1901	Additional Surgery	A070504	Premature Discharge of Battery	30/04/2021	30/04/2021
									A071102	Failure to Interrogate	30/04/2021	30/04/2021
					CT145	Cardiovascular devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	30/04/2021	30/04/2021
									A071102	Failure to Interrogate	30/04/2021	30/04/2021
							F1901	Additional Surgery	A070504	Premature Discharge of Battery	30/04/2021	30/04/2021
									A071102	Failure to Interrogate	30/04/2021	30/04/2021
					CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	30/04/2021	30/04/2021
									A071102	Failure to Interrogate	30/04/2021	30/04/2021
							F1901	Additional Surgery	A070504	Premature Discharge of Battery	30/04/2021	30/04/2021
									A071102	Failure to Interrogate	30/04/2021	30/04/2021
					CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	30/04/2021	30/04/2021
									A071102	Failure to Interrogate	30/04/2021	30/04/2021
							F1901	Additional Surgery	A070504	Premature Discharge of Battery	30/04/2021	30/04/2021
									A071102	Failure to Interrogate	30/04/2021	30/04/2021
					CT233	Surgical	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	30/04/2021	30/04/2021
									A071102	Failure to Interrogate	30/04/2021	30/04/2021
							F1901	Additional Surgery	A070504	Premature Discharge of Battery	30/04/2021	30/04/2021
									A071102	Failure to Interrogate	30/04/2021	30/04/2021
					CT2406	Implantable	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	30/04/2021	30/04/2021
									A071102	Failure to Interrogate	30/04/2021	30/04/2021
							F1901	Additional Surgery	A070504	Premature Discharge of Battery	30/04/2021	30/04/2021
									A071102	Failure to Interrogate	30/04/2021	30/04/2021
					CT2407	Active implantable	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	30/04/2021	30/04/2021
									A071102	Failure to Interrogate	30/04/2021	30/04/2021
							F1901	Additional Surgery	A070504	Premature Discharge of Battery	30/04/2021	30/04/2021
									A071102	Failure to Interrogate	30/04/2021	30/04/2021
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	30/04/2021	30/04/2021
									A071102	Failure to Interrogate	30/04/2021	30/04/2021
							F1901	Additional Surgery	A070504	Premature Discharge of Battery	30/04/2021	30/04/2021
									A071102	Failure to Interrogate	30/04/2021	30/04/2021
					CT2727	Active implant devices/ancillaries	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	30/04/2021	30/04/2021
									A071102	Failure to Interrogate	30/04/2021	30/04/2021
							F1901	Additional Surgery	A070504	Premature Discharge of Battery	30/04/2021	30/04/2021
									A071102	Failure to Interrogate	30/04/2021	30/04/2021
					CT2750	Implantable cardiac electrical stimulators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	30/04/2021	30/04/2021
									A071102	Failure to Interrogate	30/04/2021	30/04/2021
							F1901	Additional Surgery	A070504	Premature Discharge of Battery	30/04/2021	30/04/2021
									A071102	Failure to Interrogate	30/04/2021	30/04/2021
					CT2818	Catheterization/Cannulation/Fluid path devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	30/04/2021	30/04/2021
									A071102	Failure to Interrogate	30/04/2021	30/04/2021
							F1901	Additional Surgery	A070504	Premature Discharge of Battery	30/04/2021	30/04/2021
									A071102	Failure to Interrogate	30/04/2021	30/04/2021
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	30/04/2021	30/04/2021
									A071102	Failure to Interrogate	30/04/2021	30/04/2021
							F1901	Additional Surgery	A070504	Premature Discharge of Battery	30/04/2021	30/04/2021
									A071102	Failure to Interrogate	30/04/2021	30/04/2021
					CT2985	Active	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	30/04/2021	30/04/2021
									A071102	Failure to Interrogate	30/04/2021	30/04/2021
							F1901	Additional Surgery	A070504	Premature Discharge of Battery	30/04/2021	30/04/2021
									A071102	Failure to Interrogate	30/04/2021	30/04/2021
					CT3001	Implantable metal	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	30/04/2021	30/04/2021
									A071102	Failure to Interrogate	30/04/2021	30/04/2021
							F1901	Additional Surgery	A070504	Premature Discharge of Battery	30/04/2021	30/04/2021
									A071102	Failure to Interrogate	30/04/2021	30/04/2021
					CT3002	Implantable synthetic polymer	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	30/04/2021	30/04/2021
									A071102	Failure to Interrogate	30/04/2021	30/04/2021
							F1901	Additional Surgery	A070504	Premature Discharge of Battery	30/04/2021	30/04/2021
									A071102	Failure to Interrogate	30/04/2021	30/04/2021
					CT3006	Implantable inorganic material	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	30/04/2021	30/04/2021
									A071102	Failure to Interrogate	30/04/2021	30/04/2021
							F1901	Additional Surgery	A070504	Premature Discharge of Battery	30/04/2021	30/04/2021
									A071102	Failure to Interrogate	30/04/2021	30/04/2021
					CT314	Electrically powered	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	30/04/2021	30/04/2021
									A071102	Failure to Interrogate	30/04/2021	30/04/2021
							F1901	Additional Surgery	A070504	Premature Discharge of Battery	30/04/2021	30/04/2021
									A071102	Failure to Interrogate	30/04/2021	30/04/2021
					CT321	Long-term surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	30/04/2021	30/04/2021
									A071102	Failure to Interrogate	30/04/2021	30/04/2021
							F1901	Additional Surgery	A070504	Premature Discharge of Battery	30/04/2021	30/04/2021
									A071102	Failure to Interrogate	30/04/2021	30/04/2021
					CT981	Single use	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	30/04/2021	30/04/2021
									A071102	Failure to Interrogate	30/04/2021	30/04/2021
							F1901	Additional Surgery	A070504	Premature Discharge of Battery	30/04/2021	30/04/2021

									A071102	Failure to Interrogate	30/04/2021	30/04/2021
					CT983	Surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	30/04/2021	30/04/2021
									A071102	Failure to Interrogate	30/04/2021	30/04/2021
							F1901	Additional Surgery	A070504	Premature Discharge of Battery	30/04/2021	30/04/2021
									A071102	Failure to Interrogate	30/04/2021	30/04/2021
10873	2021	All other reportable incidents	47265	Dual-chamber implantable pacemaker, rate-responsive	CT141	Electrodes/Leads and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	31/03/2021	31/03/2021
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	31/03/2021	31/03/2021
							F12	Serious Injury/ Illness/ Impairment	A070504	Premature Discharge of Battery	31/03/2021	31/03/2021
							F1903	Device Explantation	A070504	Premature Discharge of Battery	31/03/2021	31/03/2021
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	31/03/2021	31/03/2021
					CT145	Cardiovascular devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	31/03/2021	31/03/2021
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	31/03/2021	31/03/2021
							F12	Serious Injury/ Illness/ Impairment	A070504	Premature Discharge of Battery	31/03/2021	31/03/2021
							F1903	Device Explantation	A070504	Premature Discharge of Battery	31/03/2021	31/03/2021
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	31/03/2021	31/03/2021
					CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	31/03/2021	31/03/2021
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	31/03/2021	31/03/2021
							F12	Serious Injury/ Illness/ Impairment	A070504	Premature Discharge of Battery	31/03/2021	31/03/2021
							F1903	Device Explantation	A070504	Premature Discharge of Battery	31/03/2021	31/03/2021
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	31/03/2021	31/03/2021
					CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	31/03/2021	31/03/2021
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	31/03/2021	31/03/2021
							F12	Serious Injury/ Illness/ Impairment	A070504	Premature Discharge of Battery	31/03/2021	31/03/2021
							F1903	Device Explantation	A070504	Premature Discharge of Battery	31/03/2021	31/03/2021
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	31/03/2021	31/03/2021
					CT233	Surgical	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	31/03/2021	31/03/2021
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	31/03/2021	31/03/2021
							F12	Serious Injury/ Illness/ Impairment	A070504	Premature Discharge of Battery	31/03/2021	31/03/2021
							F1903	Device Explantation	A070504	Premature Discharge of Battery	31/03/2021	31/03/2021
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	31/03/2021	31/03/2021
					CT2406	Implantable	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	31/03/2021	31/03/2021
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	31/03/2021	31/03/2021
							F12	Serious Injury/ Illness/ Impairment	A070504	Premature Discharge of Battery	31/03/2021	31/03/2021
							F1903	Device Explantation	A070504	Premature Discharge of Battery	31/03/2021	31/03/2021
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	31/03/2021	31/03/2021
					CT2407	Active implantable	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	31/03/2021	31/03/2021
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	31/03/2021	31/03/2021
							F12	Serious Injury/ Illness/ Impairment	A070504	Premature Discharge of Battery	31/03/2021	31/03/2021
							F1903	Device Explantation	A070504	Premature Discharge of Battery	31/03/2021	31/03/2021
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	31/03/2021	31/03/2021
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	31/03/2021	31/03/2021
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	31/03/2021	31/03/2021
							F12	Serious Injury/ Illness/ Impairment	A070504	Premature Discharge of Battery	31/03/2021	31/03/2021
							F1903	Device Explantation	A070504	Premature Discharge of Battery	31/03/2021	31/03/2021
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	31/03/2021	31/03/2021
					CT2727	Active implant devices/ancillaries	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	31/03/2021	31/03/2021
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	31/03/2021	31/03/2021
							F12	Serious Injury/ Illness/ Impairment	A070504	Premature Discharge of Battery	31/03/2021	31/03/2021
							F1903	Device Explantation	A070504	Premature Discharge of Battery	31/03/2021	31/03/2021
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	31/03/2021	31/03/2021
					CT2750	Implantable cardiac electrical stimulators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	31/03/2021	31/03/2021
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	31/03/2021	31/03/2021
							F12	Serious Injury/ Illness/ Impairment	A070504	Premature Discharge of Battery	31/03/2021	31/03/2021
							F1903	Device Explantation	A070504	Premature Discharge of Battery	31/03/2021	31/03/2021
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	31/03/2021	31/03/2021
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	31/03/2021	31/03/2021
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	31/03/2021	31/03/2021
							F12	Serious Injury/ Illness/ Impairment	A070504	Premature Discharge of Battery	31/03/2021	31/03/2021
							F1903	Device Explantation	A070504	Premature Discharge of Battery	31/03/2021	31/03/2021
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	31/03/2021	31/03/2021
					CT2985	Active	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	31/03/2021	31/03/2021
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	31/03/2021	31/03/2021
							F12	Serious Injury/ Illness/ Impairment	A070504	Premature Discharge of Battery	31/03/2021	31/03/2021
							F1903	Device Explantation	A070504	Premature Discharge of Battery	31/03/2021	31/03/2021
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	31/03/2021	31/03/2021
					CT3001	Implantable metal	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	31/03/2021	31/03/2021
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	31/03/2021	31/03/2021
							F12	Serious Injury/ Illness/ Impairment	A070504	Premature Discharge of Battery	31/03/2021	31/03/2021
							F1903	Device Explantation	A070504	Premature Discharge of Battery	31/03/2021	31/03/2021
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	31/03/2021	31/03/2021
					CT3002	Implantable synthetic polymer	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	31/03/2021	31/03/2021
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	31/03/2021	31/03/2021
							F12	Serious Injury/ Illness/ Impairment	A070504	Premature Discharge of Battery	31/03/2021	31/03/2021
							F1903	Device Explantation	A070504	Premature Discharge of Battery	31/03/2021	31/03/2021
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	31/03/2021	31/03/2021
					CT3006	Implantable inorganic material	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	31/03/2021	31/03/2021
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	31/03/2021	31/03/2021
							F12	Serious Injury/ Illness/ Impairment	A070504	Premature Discharge of Battery	31/03/2021	31/03/2021
							F1903	Device Explantation	A070504	Premature Discharge of Battery	31/03/2021	31/03/2021
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	31/03/2021	31/03/2021
					CT314	Electrically powered	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	31/03/2021	31/03/2021
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	31/03/2021	31/03/2021
							F12	Serious Injury/ Illness/ Impairment	A070504	Premature Discharge of Battery	31/03/2021	31/03/2021
							F1903	Device Explantation	A070504	Premature Discharge of Battery	31/03/2021	31/03/2021
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	31/03/2021	31/03/2021
					CT321	Long-term surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	31/03/2021	31/03/2021
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	31/03/2021	31/03/2021
							F12	Serious Injury/ Illness/ Impairment	A070504	Premature Discharge of Battery	31/03/2021	31/03/2021

							F1903	Device Explantation	A070504	Premature Discharge of Battery	31/03/2021	31/03/2021
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	31/03/2021	31/03/2021
					CT449	Generators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	31/03/2021	31/03/2021
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	31/03/2021	31/03/2021
							F12	Serious Injury/ Illness/ Impairment	A070504	Premature Discharge of Battery	31/03/2021	31/03/2021
							F1903	Device Explantation	A070504	Premature Discharge of Battery	31/03/2021	31/03/2021
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	31/03/2021	31/03/2021
					CT981	Single use	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	31/03/2021	31/03/2021
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	31/03/2021	31/03/2021
							F12	Serious Injury/ Illness/ Impairment	A070504	Premature Discharge of Battery	31/03/2021	31/03/2021
							F1903	Device Explantation	A070504	Premature Discharge of Battery	31/03/2021	31/03/2021
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	31/03/2021	31/03/2021
					CT983	Surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	31/03/2021	31/03/2021
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	31/03/2021	31/03/2021
							F12	Serious Injury/ Illness/ Impairment	A070504	Premature Discharge of Battery	31/03/2021	31/03/2021
							F1903	Device Explantation	A070504	Premature Discharge of Battery	31/03/2021	31/03/2021
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	31/03/2021	31/03/2021
10874	2021	All other reportable incidents	35852	Single-chamber implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A070909	Oversensing		
							F08	Hospitalization or Prolonged Hospitalization	A070909	Oversensing		
							F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A070909	Oversensing		
10875	2021	All other reportable incidents	37265	Dual-chamber implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	14/04/2021	14/04/2021
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	14/04/2021	14/04/2021
							F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A070504	Premature Discharge of Battery	14/04/2021	14/04/2021
							F1903	Device Explantation	A070504	Premature Discharge of Battery	14/04/2021	14/04/2021
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	14/04/2021	14/04/2021
10876	2021	All other reportable incidents	47267	Single-chamber implantable pacemaker, rate-responsive	CT141	Electrodes/Leads and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	13/04/2021	13/04/2021
							F12	Serious Injury/ Illness/ Impairment	A070504	Premature Discharge of Battery	13/04/2021	13/04/2021
							F19	Surgical Intervention	A070504	Premature Discharge of Battery	13/04/2021	13/04/2021
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	13/04/2021	13/04/2021
					CT145	Cardiovascular devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	13/04/2021	13/04/2021
							F12	Serious Injury/ Illness/ Impairment	A070504	Premature Discharge of Battery	13/04/2021	13/04/2021
							F19	Surgical Intervention	A070504	Premature Discharge of Battery	13/04/2021	13/04/2021
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	13/04/2021	13/04/2021
					CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	13/04/2021	13/04/2021
							F12	Serious Injury/ Illness/ Impairment	A070504	Premature Discharge of Battery	13/04/2021	13/04/2021
							F19	Surgical Intervention	A070504	Premature Discharge of Battery	13/04/2021	13/04/2021
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	13/04/2021	13/04/2021
					CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	13/04/2021	13/04/2021
							F12	Serious Injury/ Illness/ Impairment	A070504	Premature Discharge of Battery	13/04/2021	13/04/2021
							F19	Surgical Intervention	A070504	Premature Discharge of Battery	13/04/2021	13/04/2021
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	13/04/2021	13/04/2021
					CT233	Surgical	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	13/04/2021	13/04/2021
							F12	Serious Injury/ Illness/ Impairment	A070504	Premature Discharge of Battery	13/04/2021	13/04/2021
							F19	Surgical Intervention	A070504	Premature Discharge of Battery	13/04/2021	13/04/2021
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	13/04/2021	13/04/2021
					CT2406	Implantable	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	13/04/2021	13/04/2021
							F12	Serious Injury/ Illness/ Impairment	A070504	Premature Discharge of Battery	13/04/2021	13/04/2021
							F19	Surgical Intervention	A070504	Premature Discharge of Battery	13/04/2021	13/04/2021
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	13/04/2021	13/04/2021
					CT2407	Active implantable	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	13/04/2021	13/04/2021
							F12	Serious Injury/ Illness/ Impairment	A070504	Premature Discharge of Battery	13/04/2021	13/04/2021
							F19	Surgical Intervention	A070504	Premature Discharge of Battery	13/04/2021	13/04/2021
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	13/04/2021	13/04/2021
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	13/04/2021	13/04/2021
							F12	Serious Injury/ Illness/ Impairment	A070504	Premature Discharge of Battery	13/04/2021	13/04/2021
							F19	Surgical Intervention	A070504	Premature Discharge of Battery	13/04/2021	13/04/2021
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	13/04/2021	13/04/2021
					CT2727	Active implant devices/ancillaries	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	13/04/2021	13/04/2021
							F12	Serious Injury/ Illness/ Impairment	A070504	Premature Discharge of Battery	13/04/2021	13/04/2021
							F19	Surgical Intervention	A070504	Premature Discharge of Battery	13/04/2021	13/04/2021
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	13/04/2021	13/04/2021
					CT2750	Implantable cardiac electrical stimulators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	13/04/2021	13/04/2021
							F12	Serious Injury/ Illness/ Impairment	A070504	Premature Discharge of Battery	13/04/2021	13/04/2021
							F19	Surgical Intervention	A070504	Premature Discharge of Battery	13/04/2021	13/04/2021
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	13/04/2021	13/04/2021
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	13/04/2021	13/04/2021
							F12	Serious Injury/ Illness/ Impairment	A070504	Premature Discharge of Battery	13/04/2021	13/04/2021
							F19	Surgical Intervention	A070504	Premature Discharge of Battery	13/04/2021	13/04/2021
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	13/04/2021	13/04/2021
					CT2985	Active	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	13/04/2021	13/04/2021
							F12	Serious Injury/ Illness/ Impairment	A070504	Premature Discharge of Battery	13/04/2021	13/04/2021
							F19	Surgical Intervention	A070504	Premature Discharge of Battery	13/04/2021	13/04/2021
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	13/04/2021	13/04/2021
					CT3001	Implantable metal	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	13/04/2021	13/04/2021
							F12	Serious Injury/ Illness/ Impairment	A070504	Premature Discharge of Battery	13/04/2021	13/04/2021
							F19	Surgical Intervention	A070504	Premature Discharge of Battery	13/04/2021	13/04/2021
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	13/04/2021	13/04/2021
					CT3002	Implantable synthetic polymer	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	13/04/2021	13/04/2021
							F12	Serious Injury/ Illness/ Impairment	A070504	Premature Discharge of Battery	13/04/2021	13/04/2021
							F19	Surgical Intervention	A070504	Premature Discharge of Battery	13/04/2021	13/04/2021
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	13/04/2021	13/04/2021
					CT3006	Implantable inorganic material	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	13/04/2021	13/04/2021
							F12	Serious Injury/ Illness/ Impairment	A070504	Premature Discharge of Battery	13/04/2021	13/04/2021
							F19	Surgical Intervention	A070504	Premature Discharge of Battery	13/04/2021	13/04/2021
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	13/04/2021	13/04/2021
					CT314	Electrically powered	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	13/04/2021	13/04/2021
							F12	Serious Injury/ Illness/ Impairment	A070504	Premature Discharge of Battery	13/04/2021	13/04/2021
							F19	Surgical Intervention	A070504	Premature Discharge of Battery	13/04/2021	13/04/2021

							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	13/04/2021	13/04/2021
					CT321	Long-term surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	13/04/2021	13/04/2021
							F12	Serious Injury/ Illness/ Impairment	A070504	Premature Discharge of Battery	13/04/2021	13/04/2021
							F19	Surgical Intervention	A070504	Premature Discharge of Battery	13/04/2021	13/04/2021
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	13/04/2021	13/04/2021
					CT449	Generators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	13/04/2021	13/04/2021
							F12	Serious Injury/ Illness/ Impairment	A070504	Premature Discharge of Battery	13/04/2021	13/04/2021
							F19	Surgical Intervention	A070504	Premature Discharge of Battery	13/04/2021	13/04/2021
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	13/04/2021	13/04/2021
					CT981	Single use	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	13/04/2021	13/04/2021
							F12	Serious Injury/ Illness/ Impairment	A070504	Premature Discharge of Battery	13/04/2021	13/04/2021
							F19	Surgical Intervention	A070504	Premature Discharge of Battery	13/04/2021	13/04/2021
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	13/04/2021	13/04/2021
					CT983	Surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	13/04/2021	13/04/2021
							F12	Serious Injury/ Illness/ Impairment	A070504	Premature Discharge of Battery	13/04/2021	13/04/2021
							F19	Surgical Intervention	A070504	Premature Discharge of Battery	13/04/2021	13/04/2021
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	13/04/2021	13/04/2021
10877	2021	All other reportable incidents	47265	Dual-chamber implantable pacemaker, rate-responsive	CT141	Electrodes/Leads and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT145	Cardiovascular devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT233	Surgical	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT2406	Implantable	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT2407	Active implantable	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT2727	Active implant devices/ancillaries	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT2750	Implantable cardiac electrical stimulators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT2985	Active	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT3001	Implantable metal	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT3002	Implantable synthetic polymer	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT3006	Implantable inorganic material	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT314	Electrically powered	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT321	Long-term surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT449	Generators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT981	Single use	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT983	Surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
10878	2021	All other reportable incidents	36241	Pacemaker/defibrillator lead	CT1684	Cardiac pacemakers and associated devices	E2103	Device Overstimulation of Tissue	A072201	High Impedance		
							F15	Recognized Device or Procedural Complication	A072201	High Impedance		
10879	2021	All other reportable incidents	47265	Dual-chamber implantable pacemaker, rate-responsive	CT141	Electrodes/Leads and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem		
							F26	No Health Consequences or Impact	A0705	Battery Problem		
					CT145	Cardiovascular devices	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem		
							F26	No Health Consequences or Impact	A0705	Battery Problem		
					CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem		
							F26	No Health Consequences or Impact	A0705	Battery Problem		
					CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem		
							F26	No Health Consequences or Impact	A0705	Battery Problem		
					CT233	Surgical	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem		
							F26	No Health Consequences or Impact	A0705	Battery Problem		
					CT2406	Implantable	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem		
							F26	No Health Consequences or Impact	A0705	Battery Problem		
					CT2407	Active implantable	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem		
							F26	No Health Consequences or Impact	A0705	Battery Problem		
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem		
							F26	No Health Consequences or Impact	A0705	Battery Problem		
					CT2727	Active implant devices/ancillaries	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem		
							F26	No Health Consequences or Impact	A0705	Battery Problem		
					CT2750	Implantable cardiac electrical stimulators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem		
							F26	No Health Consequences or Impact	A0705	Battery Problem		
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem		
							F26	No Health Consequences or Impact	A0705	Battery Problem		
					CT2985	Active	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem		
							F26	No Health Consequences or Impact	A0705	Battery Problem		
					CT3001	Implantable metal	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem		
							F26	No Health Consequences or Impact	A0705	Battery Problem		
					CT3002	Implantable synthetic polymer	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem		
							F26	No Health Consequences or Impact	A0705	Battery Problem		
					CT3006	Implantable inorganic material	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem		

						F26	No Health Consequences or Impact	A0705	Battery Problem			
					CT314	Electrically powered	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem		
							F26	No Health Consequences or Impact	A0705	Battery Problem		
					CT321	Long-term surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem		
							F26	No Health Consequences or Impact	A0705	Battery Problem		
					CT449	Generators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem		
							F26	No Health Consequences or Impact	A0705	Battery Problem		
					CT981	Single use	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem		
							F26	No Health Consequences or Impact	A0705	Battery Problem		
					CT983	Surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem		
							F26	No Health Consequences or Impact	A0705	Battery Problem		
10880	2021	All other reportable incidents	47265	Dual-chamber implantable pacemaker, rate-responsive	CT141	Electrodes/Leads and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem		
							F24	Insufficient Health Impact Information	A0705	Battery Problem		
					CT145	Cardiovascular devices	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem		
							F24	Insufficient Health Impact Information	A0705	Battery Problem		
					CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem		
							F24	Insufficient Health Impact Information	A0705	Battery Problem		
					CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem		
							F24	Insufficient Health Impact Information	A0705	Battery Problem		
					CT233	Surgical	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem		
							F24	Insufficient Health Impact Information	A0705	Battery Problem		
					CT2406	Implantable	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem		
							F24	Insufficient Health Impact Information	A0705	Battery Problem		
					CT2407	Active implantable	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem		
							F24	Insufficient Health Impact Information	A0705	Battery Problem		
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem		
							F24	Insufficient Health Impact Information	A0705	Battery Problem		
					CT2727	Active implant devices/ancillaries	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem		
							F24	Insufficient Health Impact Information	A0705	Battery Problem		
					CT2750	Implantable cardiac electrical stimulators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem		
							F24	Insufficient Health Impact Information	A0705	Battery Problem		
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem		
							F24	Insufficient Health Impact Information	A0705	Battery Problem		
					CT2985	Active	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem		
							F24	Insufficient Health Impact Information	A0705	Battery Problem		
					CT3001	Implantable metal	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem		
							F24	Insufficient Health Impact Information	A0705	Battery Problem		
					CT3002	Implantable synthetic polymer	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem		
							F24	Insufficient Health Impact Information	A0705	Battery Problem		
					CT3006	Implantable inorganic material	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem		
							F24	Insufficient Health Impact Information	A0705	Battery Problem		
					CT314	Electrically powered	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem		
							F24	Insufficient Health Impact Information	A0705	Battery Problem		
					CT321	Long-term surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem		
							F24	Insufficient Health Impact Information	A0705	Battery Problem		
					CT449	Generators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem		
							F24	Insufficient Health Impact Information	A0705	Battery Problem		
					CT981	Single use	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem		
							F24	Insufficient Health Impact Information	A0705	Battery Problem		
					CT983	Surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem		
							F24	Insufficient Health Impact Information	A0705	Battery Problem		
10881	2021	All other reportable incidents	36241	Pacemaker/defibrillator lead	CT1684	Cardiac pacemakers and associated devices	E2401	Insufficient Information	A070909	Oversensing		
							E2403	No Clinical Signs, Symptoms or Conditions	A070909	Oversensing		
							F15	Recognized Device or Procedural Complication	A070909	Oversensing		
10882	2021	All other reportable incidents	47270	Cardiac resynchronization therapy implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E2401	Insufficient Information	A070101	Failure to Capture		
									A070909	Oversensing		
							E2403	No Clinical Signs, Symptoms or Conditions	A070101	Failure to Capture		
									A070909	Oversensing		
							F15	Recognized Device or Procedural Complication	A070101	Failure to Capture		
									A070909	Oversensing		
10883	2021	All other reportable incidents	35852	Single-chamber implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E2401	Insufficient Information	A070909	Oversensing		
							E2403	No Clinical Signs, Symptoms or Conditions	A070909	Oversensing		
							F26	No Health Consequences or Impact	A070909	Oversensing		
10884	2021	All other reportable incidents	35852	Single-chamber implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A070909	Oversensing		
							F26	No Health Consequences or Impact	A070909	Oversensing		
10885	2021	All other reportable incidents	47263	Cardiac resynchronization therapy implantable pacemaker	CT1684	Cardiac pacemakers and associated devices	E2103	Device Overstimulation of Tissue	A071209	Extracardiac Stimulation	03/05/2021	03/05/2021
									A0908	Incorrect, Inadequate or Imprecise Result or Readings	03/05/2021	03/05/2021
							E2403	No Clinical Signs, Symptoms or Conditions	A071209	Extracardiac Stimulation	03/05/2021	03/05/2021
									A0908	Incorrect, Inadequate or Imprecise Result or Readings	03/05/2021	03/05/2021
							F08	Hospitalization or Prolonged Hospitalization	A071209	Extracardiac Stimulation	03/05/2021	03/05/2021
									A0908	Incorrect, Inadequate or Imprecise Result or Readings	03/05/2021	03/05/2021
							F12	Serious Injury/ Illness/ Impairment	A071209	Extracardiac Stimulation	03/05/2021	03/05/2021
									A0908	Incorrect, Inadequate or Imprecise Result or Readings	03/05/2021	03/05/2021
							F19	Surgical Intervention	A071209	Extracardiac Stimulation	03/05/2021	03/05/2021
									A0908	Incorrect, Inadequate or Imprecise Result or Readings	03/05/2021	03/05/2021
					CT1685	Cardiac pacemakers	E2103	Device Overstimulation of Tissue	A071209	Extracardiac Stimulation	03/05/2021	03/05/2021
									A0908	Incorrect, Inadequate or Imprecise Result or Readings	03/05/2021	03/05/2021
							E2403	No Clinical Signs, Symptoms or Conditions	A071209	Extracardiac Stimulation	03/05/2021	03/05/2021
									A0908	Incorrect, Inadequate or Imprecise Result or Readings	03/05/2021	03/05/2021
							F08	Hospitalization or Prolonged Hospitalization	A071209	Extracardiac Stimulation	03/05/2021	03/05/2021
									A0908	Incorrect, Inadequate or Imprecise Result or Readings	03/05/2021	03/05/2021
							F12	Serious Injury/ Illness/ Impairment	A071209	Extracardiac Stimulation	03/05/2021	03/05/2021

								A0908	Incorrect, Inadequate or Imprecise Result or Readings	03/05/2021	03/05/2021
						F19	Surgical Intervention	A071209	Extracardiac Stimulation	03/05/2021	03/05/2021
								A0908	Incorrect, Inadequate or Imprecise Result or Readings	03/05/2021	03/05/2021
				CT257	Implantable cardiac pacemakers	E2103	Device Overstimulation of Tissue	A071209	Extracardiac Stimulation	03/05/2021	03/05/2021
								A0908	Incorrect, Inadequate or Imprecise Result or Readings	03/05/2021	03/05/2021
						E2403	No Clinical Signs, Symptoms or Conditions	A071209	Extracardiac Stimulation	03/05/2021	03/05/2021
								A0908	Incorrect, Inadequate or Imprecise Result or Readings	03/05/2021	03/05/2021
						F08	Hospitalization or Prolonged Hospitalization	A071209	Extracardiac Stimulation	03/05/2021	03/05/2021
								A0908	Incorrect, Inadequate or Imprecise Result or Readings	03/05/2021	03/05/2021
						F12	Serious Injury/ Illness/ Impairment	A071209	Extracardiac Stimulation	03/05/2021	03/05/2021
								A0908	Incorrect, Inadequate or Imprecise Result or Readings	03/05/2021	03/05/2021
						F19	Surgical Intervention	A071209	Extracardiac Stimulation	03/05/2021	03/05/2021
								A0908	Incorrect, Inadequate or Imprecise Result or Readings	03/05/2021	03/05/2021
				CT2866	Implantable cardiac electrical stimulators	E2103	Device Overstimulation of Tissue	A071209	Extracardiac Stimulation	03/05/2021	03/05/2021
								A0908	Incorrect, Inadequate or Imprecise Result or Readings	03/05/2021	03/05/2021
						E2403	No Clinical Signs, Symptoms or Conditions	A071209	Extracardiac Stimulation	03/05/2021	03/05/2021
								A0908	Incorrect, Inadequate or Imprecise Result or Readings	03/05/2021	03/05/2021
						F08	Hospitalization or Prolonged Hospitalization	A071209	Extracardiac Stimulation	03/05/2021	03/05/2021
								A0908	Incorrect, Inadequate or Imprecise Result or Readings	03/05/2021	03/05/2021
						F12	Serious Injury/ Illness/ Impairment	A071209	Extracardiac Stimulation	03/05/2021	03/05/2021
								A0908	Incorrect, Inadequate or Imprecise Result or Readings	03/05/2021	03/05/2021
						F19	Surgical Intervention	A071209	Extracardiac Stimulation	03/05/2021	03/05/2021
								A0908	Incorrect, Inadequate or Imprecise Result or Readings	03/05/2021	03/05/2021
10886	2021	All other reportable incidents	36241	Pacemaker/defibrillator lead	CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070909	Oversensing		
									A0712	Pacing Problem		
									A072201	High Impedance		
						F08	Hospitalization or Prolonged Hospitalization	A070909	Oversensing		
								A0712	Pacing Problem		
								A072201	High Impedance		
10887	2021	All other reportable incidents	47270	Cardiac resynchronization therapy implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A23	Use of Device Problem		
							F12	Serious Injury/ Illness/ Impairment	A23	Use of Device Problem		
							F19	Surgical Intervention	A23	Use of Device Problem		
10888	2021	All other reportable incidents	47265	Dual-chamber implantable pacemaker, rate-responsive	CT141	Electrodes/Leads and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F27	Problem Identified During Non-Clinical Procedure	A070504	Premature Discharge of Battery		
					CT145	Cardiovascular devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F27	Problem Identified During Non-Clinical Procedure	A070504	Premature Discharge of Battery		
					CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F27	Problem Identified During Non-Clinical Procedure	A070504	Premature Discharge of Battery		
					CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F27	Problem Identified During Non-Clinical Procedure	A070504	Premature Discharge of Battery		
					CT233	Surgical	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F27	Problem Identified During Non-Clinical Procedure	A070504	Premature Discharge of Battery		
					CT2406	Implantable	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F27	Problem Identified During Non-Clinical Procedure	A070504	Premature Discharge of Battery		
					CT2407	Active implantable	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F27	Problem Identified During Non-Clinical Procedure	A070504	Premature Discharge of Battery		
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F27	Problem Identified During Non-Clinical Procedure	A070504	Premature Discharge of Battery		
					CT2727	Active implant devices/ancillaries	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F27	Problem Identified During Non-Clinical Procedure	A070504	Premature Discharge of Battery		
					CT2750	Implantable cardiac electrical stimulators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F27	Problem Identified During Non-Clinical Procedure	A070504	Premature Discharge of Battery		
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F27	Problem Identified During Non-Clinical Procedure	A070504	Premature Discharge of Battery		
					CT2985	Active	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F27	Problem Identified During Non-Clinical Procedure	A070504	Premature Discharge of Battery		
					CT3001	Implantable metal	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F27	Problem Identified During Non-Clinical Procedure	A070504	Premature Discharge of Battery		
					CT3002	Implantable synthetic polymer	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F27	Problem Identified During Non-Clinical Procedure	A070504	Premature Discharge of Battery		
					CT3006	Implantable inorganic material	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F27	Problem Identified During Non-Clinical Procedure	A070504	Premature Discharge of Battery		
					CT314	Electrically powered	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F27	Problem Identified During Non-Clinical Procedure	A070504	Premature Discharge of Battery		
					CT321	Long-term surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F27	Problem Identified During Non-Clinical Procedure	A070504	Premature Discharge of Battery		
					CT449	Generators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F27	Problem Identified During Non-Clinical Procedure	A070504	Premature Discharge of Battery		
					CT981	Single use	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F27	Problem Identified During Non-Clinical Procedure	A070504	Premature Discharge of Battery		
					CT983	Surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F27	Problem Identified During Non-Clinical Procedure	A070504	Premature Discharge of Battery		
10889	2021	All other reportable incidents	35852	Single-chamber implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A1301	Failure to Read Input Signal		
							F23	Unexpected Medical Intervention	A1301	Failure to Read Input Signal		
10890	2021	All other reportable incidents	35852	Single-chamber implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	01/04/2021	01/04/2021
							F1901	Additional Surgery	A070504	Premature Discharge of Battery	01/04/2021	01/04/2021
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	01/04/2021	01/04/2021
10891	2021	All other reportable incidents	47265	Dual-chamber implantable pacemaker, rate-responsive	CT141	Electrodes/Leads and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		

					CT145	Cardiovascular devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT233	Surgical	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT2406	Implantable	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT2407	Active implantable	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT2727	Active implant devices/ancillaries	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT2750	Implantable cardiac electrical stimulators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT2985	Active	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT3001	Implantable metal	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT3002	Implantable synthetic polymer	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT3006	Implantable inorganic material	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT314	Electrically powered	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT321	Long-term surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT449	Generators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT981	Single use	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT983	Surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
10892	2021	All other reportable incidents	36241	Pacemaker/defibrillator lead	CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070909	Oversensing		
							F26	No Health Consequences or Impact	A070909	Oversensing		
10893	2021	All other reportable incidents	47265	Dual-chamber implantable pacemaker, rate-responsive	CT141	Electrodes/Leads and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
									A072201	High Impedance		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
									A072201	High Impedance		
					CT145	Cardiovascular devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
									A072201	High Impedance		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
									A072201	High Impedance		
					CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
									A072201	High Impedance		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
									A072201	High Impedance		
					CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
									A072201	High Impedance		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
									A072201	High Impedance		
					CT233	Surgical	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
									A072201	High Impedance		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
									A072201	High Impedance		
					CT2406	Implantable	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
									A072201	High Impedance		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
									A072201	High Impedance		
					CT2407	Active implantable	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
									A072201	High Impedance		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
									A072201	High Impedance		
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
									A072201	High Impedance		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
									A072201	High Impedance		
					CT2727	Active implant devices/ancillaries	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
									A072201	High Impedance		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
									A072201	High Impedance		
					CT2750	Implantable cardiac electrical stimulators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
									A072201	High Impedance		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
									A072201	High Impedance		
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
									A072201	High Impedance		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
									A072201	High Impedance		
					CT2985	Active	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
									A072201	High Impedance		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
									A072201	High Impedance		

					CT3001	Implantable metal	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
									A072201	High Impedance		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
									A072201	High Impedance		
					CT3002	Implantable synthetic polymer	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
									A072201	High Impedance		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
									A072201	High Impedance		
					CT3006	Implantable inorganic material	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
									A072201	High Impedance		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
									A072201	High Impedance		
					CT314	Electrically powered	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
									A072201	High Impedance		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
									A072201	High Impedance		
					CT321	Long-term surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
									A072201	High Impedance		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
									A072201	High Impedance		
					CT449	Generators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
									A072201	High Impedance		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
									A072201	High Impedance		
					CT981	Single use	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
									A072201	High Impedance		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
									A072201	High Impedance		
					CT983	Surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
									A072201	High Impedance		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
									A072201	High Impedance		
10894	2021	All other reportable incidents	35852	Single-chamber implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem	12/04/2021	12/04/2021
									A070504	Premature Discharge of Battery	12/04/2021	12/04/2021
							F08	Hospitalization or Prolonged Hospitalization	A0705	Battery Problem	12/04/2021	12/04/2021
									A070504	Premature Discharge of Battery	12/04/2021	12/04/2021
							F1905	Device Revision or Replacement	A0705	Battery Problem	12/04/2021	12/04/2021
									A070504	Premature Discharge of Battery	12/04/2021	12/04/2021
10895	2021	All other reportable incidents	47263	Cardiac resynchronization therapy implantable pacemaker	CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070909	Oversensing		
							F26	No Health Consequences or Impact	A070909	Oversensing		
					CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A070909	Oversensing		
							F26	No Health Consequences or Impact	A070909	Oversensing		
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A070909	Oversensing		
							F26	No Health Consequences or Impact	A070909	Oversensing		
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A070909	Oversensing		
							F26	No Health Consequences or Impact	A070909	Oversensing		
10896	2021	All other reportable incidents	47263	Cardiac resynchronization therapy implantable pacemaker	CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A0908	Incorrect, Inadequate or Imprecise Result or Readings	01/04/2021	01/04/2021
							F08	Hospitalization or Prolonged Hospitalization	A0908	Incorrect, Inadequate or Imprecise Result or Readings	01/04/2021	01/04/2021
							F12	Serious Injury/ Illness/ Impairment	A0908	Incorrect, Inadequate or Imprecise Result or Readings	01/04/2021	01/04/2021
							F19	Surgical Intervention	A0908	Incorrect, Inadequate or Imprecise Result or Readings	01/04/2021	01/04/2021
							F1905	Device Revision or Replacement	A0908	Incorrect, Inadequate or Imprecise Result or Readings	01/04/2021	01/04/2021
					CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A0908	Incorrect, Inadequate or Imprecise Result or Readings	01/04/2021	01/04/2021
							F08	Hospitalization or Prolonged Hospitalization	A0908	Incorrect, Inadequate or Imprecise Result or Readings	01/04/2021	01/04/2021
							F12	Serious Injury/ Illness/ Impairment	A0908	Incorrect, Inadequate or Imprecise Result or Readings	01/04/2021	01/04/2021
							F19	Surgical Intervention	A0908	Incorrect, Inadequate or Imprecise Result or Readings	01/04/2021	01/04/2021
							F1905	Device Revision or Replacement	A0908	Incorrect, Inadequate or Imprecise Result or Readings	01/04/2021	01/04/2021
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A0908	Incorrect, Inadequate or Imprecise Result or Readings	01/04/2021	01/04/2021
							F08	Hospitalization or Prolonged Hospitalization	A0908	Incorrect, Inadequate or Imprecise Result or Readings	01/04/2021	01/04/2021
							F12	Serious Injury/ Illness/ Impairment	A0908	Incorrect, Inadequate or Imprecise Result or Readings	01/04/2021	01/04/2021
							F19	Surgical Intervention	A0908	Incorrect, Inadequate or Imprecise Result or Readings	01/04/2021	01/04/2021
							F1905	Device Revision or Replacement	A0908	Incorrect, Inadequate or Imprecise Result or Readings	01/04/2021	01/04/2021
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A0908	Incorrect, Inadequate or Imprecise Result or Readings	01/04/2021	01/04/2021
							F08	Hospitalization or Prolonged Hospitalization	A0908	Incorrect, Inadequate or Imprecise Result or Readings	01/04/2021	01/04/2021
							F12	Serious Injury/ Illness/ Impairment	A0908	Incorrect, Inadequate or Imprecise Result or Readings	01/04/2021	01/04/2021
							F19	Surgical Intervention	A0908	Incorrect, Inadequate or Imprecise Result or Readings	01/04/2021	01/04/2021
							F1905	Device Revision or Replacement	A0908	Incorrect, Inadequate or Imprecise Result or Readings	01/04/2021	01/04/2021
10897	2021	All other reportable incidents	35852	Single-chamber implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem	20/04/2021	20/04/2021
									A070504	Premature Discharge of Battery	20/04/2021	20/04/2021
							F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A0705	Battery Problem	20/04/2021	20/04/2021
									A070504	Premature Discharge of Battery	20/04/2021	20/04/2021
							F19	Surgical Intervention	A0705	Battery Problem	20/04/2021	20/04/2021
									A070504	Premature Discharge of Battery	20/04/2021	20/04/2021

10898	2021	All other reportable incidents	47265	Dual-chamber implantable pacemaker, rate-responsive	CT141	Electrodes/Leads and associated devices	E2401	Insufficient Information	A0705	Battery Problem		
							E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem		
							F15	Recognized Device or Procedural Complication	A0705	Battery Problem		
					CT145	Cardiovascular devices	E2401	Insufficient Information	A0705	Battery Problem		
							E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem		
							F15	Recognized Device or Procedural Complication	A0705	Battery Problem		
					CT1684	Cardiac pacemakers and associated devices	E2401	Insufficient Information	A0705	Battery Problem		
							E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem		
							F15	Recognized Device or Procedural Complication	A0705	Battery Problem		
					CT1685	Cardiac pacemakers	E2401	Insufficient Information	A0705	Battery Problem		
							E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem		
							F15	Recognized Device or Procedural Complication	A0705	Battery Problem		
					CT233	Surgical	E2401	Insufficient Information	A0705	Battery Problem		
							E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem		
							F15	Recognized Device or Procedural Complication	A0705	Battery Problem		
					CT2406	Implantable	E2401	Insufficient Information	A0705	Battery Problem		
							E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem		
							F15	Recognized Device or Procedural Complication	A0705	Battery Problem		
					CT2407	Active implantable	E2401	Insufficient Information	A0705	Battery Problem		
							E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem		
							F15	Recognized Device or Procedural Complication	A0705	Battery Problem		
					CT257	Implantable cardiac pacemakers	E2401	Insufficient Information	A0705	Battery Problem		
							E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem		
							F15	Recognized Device or Procedural Complication	A0705	Battery Problem		
					CT2727	Active implant devices/ancillaries	E2401	Insufficient Information	A0705	Battery Problem		
							E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem		
							F15	Recognized Device or Procedural Complication	A0705	Battery Problem		
					CT2750	Implantable cardiac electrical stimulators and associated devices	E2401	Insufficient Information	A0705	Battery Problem		
							E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem		
							F15	Recognized Device or Procedural Complication	A0705	Battery Problem		
					CT2866	Implantable cardiac electrical stimulators	E2401	Insufficient Information	A0705	Battery Problem		
							E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem		
							F15	Recognized Device or Procedural Complication	A0705	Battery Problem		
					CT2985	Active	E2401	Insufficient Information	A0705	Battery Problem		
							E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem		
							F15	Recognized Device or Procedural Complication	A0705	Battery Problem		
					CT3001	Implantable metal	E2401	Insufficient Information	A0705	Battery Problem		
							E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem		
							F15	Recognized Device or Procedural Complication	A0705	Battery Problem		
					CT3002	Implantable synthetic polymer	E2401	Insufficient Information	A0705	Battery Problem		
							E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem		
							F15	Recognized Device or Procedural Complication	A0705	Battery Problem		
					CT3006	Implantable inorganic material	E2401	Insufficient Information	A0705	Battery Problem		
							E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem		
							F15	Recognized Device or Procedural Complication	A0705	Battery Problem		
					CT314	Electrically powered	E2401	Insufficient Information	A0705	Battery Problem		
							E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem		
							F15	Recognized Device or Procedural Complication	A0705	Battery Problem		
					CT321	Long-term surgical invasive	E2401	Insufficient Information	A0705	Battery Problem		
							E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem		
							F15	Recognized Device or Procedural Complication	A0705	Battery Problem		
					CT449	Generators and associated devices	E2401	Insufficient Information	A0705	Battery Problem		
							E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem		
							F15	Recognized Device or Procedural Complication	A0705	Battery Problem		
					CT981	Single use	E2401	Insufficient Information	A0705	Battery Problem		
							E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem		
							F15	Recognized Device or Procedural Complication	A0705	Battery Problem		
					CT983	Surgical invasive	E2401	Insufficient Information	A0705	Battery Problem		
							E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem		
							F15	Recognized Device or Procedural Complication	A0705	Battery Problem		
10899	2021	All other reportable incidents	47265	Dual-chamber implantable pacemaker, rate-responsive	CT141	Electrodes/Leads and associated devices	E2401	Insufficient Information	A070101	Failure to Capture		
							E2403	No Clinical Signs, Symptoms or Conditions	A070101	Failure to Capture		
							F26	No Health Consequences or Impact	A070101	Failure to Capture		
					CT145	Cardiovascular devices	E2401	Insufficient Information	A070101	Failure to Capture		
							E2403	No Clinical Signs, Symptoms or Conditions	A070101	Failure to Capture		
							F26	No Health Consequences or Impact	A070101	Failure to Capture		
					CT1684	Cardiac pacemakers and associated devices	E2401	Insufficient Information	A070101	Failure to Capture		
							E2403	No Clinical Signs, Symptoms or Conditions	A070101	Failure to Capture		
							F26	No Health Consequences or Impact	A070101	Failure to Capture		
					CT1685	Cardiac pacemakers	E2401	Insufficient Information	A070101	Failure to Capture		
							E2403	No Clinical Signs, Symptoms or Conditions	A070101	Failure to Capture		
							F26	No Health Consequences or Impact	A070101	Failure to Capture		
					CT233	Surgical	E2401	Insufficient Information	A070101	Failure to Capture		
							E2403	No Clinical Signs, Symptoms or Conditions	A070101	Failure to Capture		
							F26	No Health Consequences or Impact	A070101	Failure to Capture		
					CT2406	Implantable	E2401	Insufficient Information	A070101	Failure to Capture		
							E2403	No Clinical Signs, Symptoms or Conditions	A070101	Failure to Capture		
							F26	No Health Consequences or Impact	A070101	Failure to Capture		
					CT2407	Active implantable	E2401	Insufficient Information	A070101	Failure to Capture		
							E2403	No Clinical Signs, Symptoms or Conditions	A070101	Failure to Capture		
							F26	No Health Consequences or Impact	A070101	Failure to Capture		
					CT257	Implantable cardiac pacemakers	E2401	Insufficient Information	A070101	Failure to Capture		
							E2403	No Clinical Signs, Symptoms or Conditions	A070101	Failure to Capture		
							F26	No Health Consequences or Impact	A070101	Failure to Capture		
					CT2727	Active implant devices/ancillaries	E2401	Insufficient Information	A070101	Failure to Capture		
							E2403	No Clinical Signs, Symptoms or Conditions	A070101	Failure to Capture		
							F26	No Health Consequences or Impact	A070101	Failure to Capture		
					CT2750	Implantable cardiac electrical stimulators and associated devices	E2401	Insufficient Information	A070101	Failure to Capture		

							E2403	No Clinical Signs, Symptoms or Conditions	A070101	Failure to Capture		
							F26	No Health Consequences or Impact	A070101	Failure to Capture		
					CT2866	Implantable cardiac electrical stimulators	E2401	Insufficient Information	A070101	Failure to Capture		
							E2403	No Clinical Signs, Symptoms or Conditions	A070101	Failure to Capture		
							F26	No Health Consequences or Impact	A070101	Failure to Capture		
					CT2985	Active	E2401	Insufficient Information	A070101	Failure to Capture		
							E2403	No Clinical Signs, Symptoms or Conditions	A070101	Failure to Capture		
							F26	No Health Consequences or Impact	A070101	Failure to Capture		
					CT3001	Implantable metal	E2401	Insufficient Information	A070101	Failure to Capture		
							E2403	No Clinical Signs, Symptoms or Conditions	A070101	Failure to Capture		
							F26	No Health Consequences or Impact	A070101	Failure to Capture		
					CT3002	Implantable synthetic polymer	E2401	Insufficient Information	A070101	Failure to Capture		
							E2403	No Clinical Signs, Symptoms or Conditions	A070101	Failure to Capture		
							F26	No Health Consequences or Impact	A070101	Failure to Capture		
					CT3006	Implantable inorganic material	E2401	Insufficient Information	A070101	Failure to Capture		
							E2403	No Clinical Signs, Symptoms or Conditions	A070101	Failure to Capture		
							F26	No Health Consequences or Impact	A070101	Failure to Capture		
					CT314	Electrically powered	E2401	Insufficient Information	A070101	Failure to Capture		
							E2403	No Clinical Signs, Symptoms or Conditions	A070101	Failure to Capture		
							F26	No Health Consequences or Impact	A070101	Failure to Capture		
					CT321	Long-term surgical invasive	E2401	Insufficient Information	A070101	Failure to Capture		
							E2403	No Clinical Signs, Symptoms or Conditions	A070101	Failure to Capture		
							F26	No Health Consequences or Impact	A070101	Failure to Capture		
					CT449	Generators and associated devices	E2401	Insufficient Information	A070101	Failure to Capture		
							E2403	No Clinical Signs, Symptoms or Conditions	A070101	Failure to Capture		
							F26	No Health Consequences or Impact	A070101	Failure to Capture		
					CT981	Single use	E2401	Insufficient Information	A070101	Failure to Capture		
							E2403	No Clinical Signs, Symptoms or Conditions	A070101	Failure to Capture		
							F26	No Health Consequences or Impact	A070101	Failure to Capture		
					CT983	Surgical invasive	E2401	Insufficient Information	A070101	Failure to Capture		
							E2403	No Clinical Signs, Symptoms or Conditions	A070101	Failure to Capture		
							F26	No Health Consequences or Impact	A070101	Failure to Capture		
10900	2021	All other reportable incidents	35852	Single-chamber implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E2104	Electric Shock	A070909	Oversensing		
									A071302	Inappropriate/Inadequate Shock/Stimulation		
							F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A070909	Oversensing		
									A071302	Inappropriate/Inadequate Shock/Stimulation		
							F26	No Health Consequences or Impact	A070909	Oversensing		
									A071302	Inappropriate/Inadequate Shock/Stimulation		
10901	2021	All other reportable incidents	35852	Single-chamber implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	01/07/2021	01/07/2021
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	01/07/2021	01/07/2021
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	01/07/2021	01/07/2021
10902	2021	All other reportable incidents	36241	Pacemaker/defibrillator lead	CT1684	Cardiac pacemakers and associated devices	E060101	Asystole	A070909	Oversensing		
									A0712	Pacing Problem		
							E2103	Device Overstimulation of Tissue	A070909	Oversensing		
									A0712	Pacing Problem		
							E2403	No Clinical Signs, Symptoms or Conditions	A070909	Oversensing		
									A0712	Pacing Problem		
							F26	No Health Consequences or Impact	A070909	Oversensing		
									A0712	Pacing Problem		
10903	2021	All other reportable incidents	35852	Single-chamber implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A160601	Premature Elective Replacement Indicator		
							F1905	Device Revision or Replacement	A160601	Premature Elective Replacement Indicator		
10904	2021	All other reportable incidents	47270	Cardiac resynchronization therapy implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	07/04/2021	07/04/2021
							F1901	Additional Surgery	A070504	Premature Discharge of Battery	07/04/2021	07/04/2021
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	07/04/2021	07/04/2021
10905	2021	All other reportable incidents	60789	Intracardiac pacemaker	CT130	Catheters and associated devices	E060104	Bradycardia	A070101	Failure to Capture	07/04/2021	07/04/2021
									A070102	High Capture Threshold	07/04/2021	07/04/2021
							F12	Serious Injury/ Illness/ Impairment	A070101	Failure to Capture	07/04/2021	07/04/2021
									A070102	High Capture Threshold	07/04/2021	07/04/2021
							F1905	Device Revision or Replacement	A070101	Failure to Capture	07/04/2021	07/04/2021
									A070102	High Capture Threshold	07/04/2021	07/04/2021
					CT145	Cardiovascular devices	E060104	Bradycardia	A070101	Failure to Capture	07/04/2021	07/04/2021
									A070102	High Capture Threshold	07/04/2021	07/04/2021
							F12	Serious Injury/ Illness/ Impairment	A070101	Failure to Capture	07/04/2021	07/04/2021
									A070102	High Capture Threshold	07/04/2021	07/04/2021
							F1905	Device Revision or Replacement	A070101	Failure to Capture	07/04/2021	07/04/2021
									A070102	High Capture Threshold	07/04/2021	07/04/2021
					CT1684	Cardiac pacemakers and associated devices	E060104	Bradycardia	A070101	Failure to Capture	07/04/2021	07/04/2021
									A070102	High Capture Threshold	07/04/2021	07/04/2021
							F12	Serious Injury/ Illness/ Impairment	A070101	Failure to Capture	07/04/2021	07/04/2021
									A070102	High Capture Threshold	07/04/2021	07/04/2021
							F1905	Device Revision or Replacement	A070101	Failure to Capture	07/04/2021	07/04/2021
									A070102	High Capture Threshold	07/04/2021	07/04/2021
					CT1685	Cardiac pacemakers	E060104	Bradycardia	A070101	Failure to Capture	07/04/2021	07/04/2021
									A070102	High Capture Threshold	07/04/2021	07/04/2021
							F12	Serious Injury/ Illness/ Impairment	A070101	Failure to Capture	07/04/2021	07/04/2021
									A070102	High Capture Threshold	07/04/2021	07/04/2021
							F1905	Device Revision or Replacement	A070101	Failure to Capture	07/04/2021	07/04/2021
									A070102	High Capture Threshold	07/04/2021	07/04/2021
					CT233	Surgical	E060104	Bradycardia	A070101	Failure to Capture	07/04/2021	07/04/2021
									A070102	High Capture Threshold	07/04/2021	07/04/2021
							F12	Serious Injury/ Illness/ Impairment	A070101	Failure to Capture	07/04/2021	07/04/2021
									A070102	High Capture Threshold	07/04/2021	07/04/2021
							F1905	Device Revision or Replacement	A070101	Failure to Capture	07/04/2021	07/04/2021
									A070102	High Capture Threshold	07/04/2021	07/04/2021
					CT2406	Implantable	E060104	Bradycardia	A070101	Failure to Capture	07/04/2021	07/04/2021
									A070102	High Capture Threshold	07/04/2021	07/04/2021
							F12	Serious Injury/ Illness/ Impairment	A070101	Failure to Capture	07/04/2021	07/04/2021
									A070102	High Capture Threshold	07/04/2021	07/04/2021

							F1905	Device Revision or Replacement	A070101	Failure to Capture	07/04/2021	07/04/2021
									A070102	High Capture Threshold	07/04/2021	07/04/2021
					CT2407	Active implantable	E060104	Bradycardia	A070101	Failure to Capture	07/04/2021	07/04/2021
									A070102	High Capture Threshold	07/04/2021	07/04/2021
							F12	Serious Injury/ Illness/ Impairment	A070101	Failure to Capture	07/04/2021	07/04/2021
									A070102	High Capture Threshold	07/04/2021	07/04/2021
							F1905	Device Revision or Replacement	A070101	Failure to Capture	07/04/2021	07/04/2021
									A070102	High Capture Threshold	07/04/2021	07/04/2021
					CT257	Implantable cardiac pacemakers	E060104	Bradycardia	A070101	Failure to Capture	07/04/2021	07/04/2021
									A070102	High Capture Threshold	07/04/2021	07/04/2021
							F12	Serious Injury/ Illness/ Impairment	A070101	Failure to Capture	07/04/2021	07/04/2021
									A070102	High Capture Threshold	07/04/2021	07/04/2021
							F1905	Device Revision or Replacement	A070101	Failure to Capture	07/04/2021	07/04/2021
									A070102	High Capture Threshold	07/04/2021	07/04/2021
					CT2727	Active implant devices/ancillaries	E060104	Bradycardia	A070101	Failure to Capture	07/04/2021	07/04/2021
									A070102	High Capture Threshold	07/04/2021	07/04/2021
							F12	Serious Injury/ Illness/ Impairment	A070101	Failure to Capture	07/04/2021	07/04/2021
									A070102	High Capture Threshold	07/04/2021	07/04/2021
							F1905	Device Revision or Replacement	A070101	Failure to Capture	07/04/2021	07/04/2021
									A070102	High Capture Threshold	07/04/2021	07/04/2021
					CT2750	Implantable cardiac electrical stimulators and associated devices	E060104	Bradycardia	A070101	Failure to Capture	07/04/2021	07/04/2021
									A070102	High Capture Threshold	07/04/2021	07/04/2021
							F12	Serious Injury/ Illness/ Impairment	A070101	Failure to Capture	07/04/2021	07/04/2021
									A070102	High Capture Threshold	07/04/2021	07/04/2021
							F1905	Device Revision or Replacement	A070101	Failure to Capture	07/04/2021	07/04/2021
									A070102	High Capture Threshold	07/04/2021	07/04/2021
					CT2818	Catheterization/Cannulation/Fluid path devices	E060104	Bradycardia	A070101	Failure to Capture	07/04/2021	07/04/2021
									A070102	High Capture Threshold	07/04/2021	07/04/2021
							F12	Serious Injury/ Illness/ Impairment	A070101	Failure to Capture	07/04/2021	07/04/2021
									A070102	High Capture Threshold	07/04/2021	07/04/2021
							F1905	Device Revision or Replacement	A070101	Failure to Capture	07/04/2021	07/04/2021
									A070102	High Capture Threshold	07/04/2021	07/04/2021
					CT2866	Implantable cardiac electrical stimulators	E060104	Bradycardia	A070101	Failure to Capture	07/04/2021	07/04/2021
									A070102	High Capture Threshold	07/04/2021	07/04/2021
							F12	Serious Injury/ Illness/ Impairment	A070101	Failure to Capture	07/04/2021	07/04/2021
									A070102	High Capture Threshold	07/04/2021	07/04/2021
							F1905	Device Revision or Replacement	A070101	Failure to Capture	07/04/2021	07/04/2021
									A070102	High Capture Threshold	07/04/2021	07/04/2021
					CT2985	Active	E060104	Bradycardia	A070101	Failure to Capture	07/04/2021	07/04/2021
									A070102	High Capture Threshold	07/04/2021	07/04/2021
							F12	Serious Injury/ Illness/ Impairment	A070101	Failure to Capture	07/04/2021	07/04/2021
									A070102	High Capture Threshold	07/04/2021	07/04/2021
							F1905	Device Revision or Replacement	A070101	Failure to Capture	07/04/2021	07/04/2021
									A070102	High Capture Threshold	07/04/2021	07/04/2021
					CT3001	Implantable metal	E060104	Bradycardia	A070101	Failure to Capture	07/04/2021	07/04/2021
									A070102	High Capture Threshold	07/04/2021	07/04/2021
							F12	Serious Injury/ Illness/ Impairment	A070101	Failure to Capture	07/04/2021	07/04/2021
									A070102	High Capture Threshold	07/04/2021	07/04/2021
							F1905	Device Revision or Replacement	A070101	Failure to Capture	07/04/2021	07/04/2021
									A070102	High Capture Threshold	07/04/2021	07/04/2021
					CT3002	Implantable synthetic polymer	E060104	Bradycardia	A070101	Failure to Capture	07/04/2021	07/04/2021
									A070102	High Capture Threshold	07/04/2021	07/04/2021
							F12	Serious Injury/ Illness/ Impairment	A070101	Failure to Capture	07/04/2021	07/04/2021
									A070102	High Capture Threshold	07/04/2021	07/04/2021
							F1905	Device Revision or Replacement	A070101	Failure to Capture	07/04/2021	07/04/2021
									A070102	High Capture Threshold	07/04/2021	07/04/2021
					CT3006	Implantable inorganic material	E060104	Bradycardia	A070101	Failure to Capture	07/04/2021	07/04/2021
									A070102	High Capture Threshold	07/04/2021	07/04/2021
							F12	Serious Injury/ Illness/ Impairment	A070101	Failure to Capture	07/04/2021	07/04/2021
									A070102	High Capture Threshold	07/04/2021	07/04/2021
							F1905	Device Revision or Replacement	A070101	Failure to Capture	07/04/2021	07/04/2021
									A070102	High Capture Threshold	07/04/2021	07/04/2021
					CT314	Electrically powered	E060104	Bradycardia	A070101	Failure to Capture	07/04/2021	07/04/2021
									A070102	High Capture Threshold	07/04/2021	07/04/2021
							F12	Serious Injury/ Illness/ Impairment	A070101	Failure to Capture	07/04/2021	07/04/2021
									A070102	High Capture Threshold	07/04/2021	07/04/2021
							F1905	Device Revision or Replacement	A070101	Failure to Capture	07/04/2021	07/04/2021
									A070102	High Capture Threshold	07/04/2021	07/04/2021
					CT321	Long-term surgical invasive	E060104	Bradycardia	A070101	Failure to Capture	07/04/2021	07/04/2021
									A070102	High Capture Threshold	07/04/2021	07/04/2021
							F12	Serious Injury/ Illness/ Impairment	A070101	Failure to Capture	07/04/2021	07/04/2021
									A070102	High Capture Threshold	07/04/2021	07/04/2021
							F1905	Device Revision or Replacement	A070101	Failure to Capture	07/04/2021	07/04/2021
									A070102	High Capture Threshold	07/04/2021	07/04/2021
					CT981	Single use	E060104	Bradycardia	A070101	Failure to Capture	07/04/2021	07/04/2021
									A070102	High Capture Threshold	07/04/2021	07/04/2021
							F12	Serious Injury/ Illness/ Impairment	A070101	Failure to Capture	07/04/2021	07/04/2021
									A070102	High Capture Threshold	07/04/2021	07/04/2021
							F1905	Device Revision or Replacement	A070101	Failure to Capture	07/04/2021	07/04/2021
									A070102	High Capture Threshold	07/04/2021	07/04/2021
					CT983	Surgical invasive	E060104	Bradycardia	A070101	Failure to Capture	07/04/2021	07/04/2021
									A070102	High Capture Threshold	07/04/2021	07/04/2021
							F12	Serious Injury/ Illness/ Impairment	A070101	Failure to Capture	07/04/2021	07/04/2021
									A070102	High Capture Threshold	07/04/2021	07/04/2021
							F1905	Device Revision or Replacement	A070101	Failure to Capture	07/04/2021	07/04/2021
									A070102	High Capture Threshold	07/04/2021	07/04/2021
10906	2021	All other reportable incidents	17882	Physiologic-monitoring defibrillation system	CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070803	Failure to Power Up		
							F27	Problem Identified During Non-Clinical Procedure	A070803	Failure to Power Up		

10907	2021	All other reportable incidents	36241	Pacemaker/defibrillator lead	CT1684	Cardiac pacemakers and associated devices	E2103	Device Overstimulation of Tissue	A070102	High Capture Threshold		
									A072201	High Impedance		
							E2403	No Clinical Signs, Symptoms or Conditions	A070102	High Capture Threshold		
									A072201	High Impedance		
							F26	No Health Consequences or Impact	A070102	High Capture Threshold		
									A072201	High Impedance		
10908	2021	All other reportable incidents	47265	Dual-chamber implantable pacemaker, rate-responsive	CT141	Electrodes/Leads and associated devices	E2330	Pain	A24	No Apparent Device Problem	25/01/2021	25/01/2021
							F08	Hospitalization or Prolonged Hospitalization	A24	No Apparent Device Problem	25/01/2021	25/01/2021
							F12	Serious Injury/ Illness/ Impairment	A24	No Apparent Device Problem	25/01/2021	25/01/2021
							F19	Surgical Intervention	A24	No Apparent Device Problem	25/01/2021	25/01/2021
							F1905	Device Revision or Replacement	A24	No Apparent Device Problem	25/01/2021	25/01/2021
					CT145	Cardiovascular devices	E2330	Pain	A24	No Apparent Device Problem	25/01/2021	25/01/2021
							F08	Hospitalization or Prolonged Hospitalization	A24	No Apparent Device Problem	25/01/2021	25/01/2021
							F12	Serious Injury/ Illness/ Impairment	A24	No Apparent Device Problem	25/01/2021	25/01/2021
							F19	Surgical Intervention	A24	No Apparent Device Problem	25/01/2021	25/01/2021
							F1905	Device Revision or Replacement	A24	No Apparent Device Problem	25/01/2021	25/01/2021
					CT1684	Cardiac pacemakers and associated devices	E2330	Pain	A24	No Apparent Device Problem	25/01/2021	25/01/2021
							F08	Hospitalization or Prolonged Hospitalization	A24	No Apparent Device Problem	25/01/2021	25/01/2021
							F12	Serious Injury/ Illness/ Impairment	A24	No Apparent Device Problem	25/01/2021	25/01/2021
							F19	Surgical Intervention	A24	No Apparent Device Problem	25/01/2021	25/01/2021
							F1905	Device Revision or Replacement	A24	No Apparent Device Problem	25/01/2021	25/01/2021
					CT1685	Cardiac pacemakers	E2330	Pain	A24	No Apparent Device Problem	25/01/2021	25/01/2021
							F08	Hospitalization or Prolonged Hospitalization	A24	No Apparent Device Problem	25/01/2021	25/01/2021
							F12	Serious Injury/ Illness/ Impairment	A24	No Apparent Device Problem	25/01/2021	25/01/2021
							F19	Surgical Intervention	A24	No Apparent Device Problem	25/01/2021	25/01/2021
							F1905	Device Revision or Replacement	A24	No Apparent Device Problem	25/01/2021	25/01/2021
					CT233	Surgical	E2330	Pain	A24	No Apparent Device Problem	25/01/2021	25/01/2021
							F08	Hospitalization or Prolonged Hospitalization	A24	No Apparent Device Problem	25/01/2021	25/01/2021
							F12	Serious Injury/ Illness/ Impairment	A24	No Apparent Device Problem	25/01/2021	25/01/2021
							F19	Surgical Intervention	A24	No Apparent Device Problem	25/01/2021	25/01/2021
							F1905	Device Revision or Replacement	A24	No Apparent Device Problem	25/01/2021	25/01/2021
					CT2406	Implantable	E2330	Pain	A24	No Apparent Device Problem	25/01/2021	25/01/2021
							F08	Hospitalization or Prolonged Hospitalization	A24	No Apparent Device Problem	25/01/2021	25/01/2021
							F12	Serious Injury/ Illness/ Impairment	A24	No Apparent Device Problem	25/01/2021	25/01/2021
							F19	Surgical Intervention	A24	No Apparent Device Problem	25/01/2021	25/01/2021
							F1905	Device Revision or Replacement	A24	No Apparent Device Problem	25/01/2021	25/01/2021
					CT2407	Active implantable	E2330	Pain	A24	No Apparent Device Problem	25/01/2021	25/01/2021
							F08	Hospitalization or Prolonged Hospitalization	A24	No Apparent Device Problem	25/01/2021	25/01/2021
							F12	Serious Injury/ Illness/ Impairment	A24	No Apparent Device Problem	25/01/2021	25/01/2021
							F19	Surgical Intervention	A24	No Apparent Device Problem	25/01/2021	25/01/2021
							F1905	Device Revision or Replacement	A24	No Apparent Device Problem	25/01/2021	25/01/2021
					CT257	Implantable cardiac pacemakers	E2330	Pain	A24	No Apparent Device Problem	25/01/2021	25/01/2021
							F08	Hospitalization or Prolonged Hospitalization	A24	No Apparent Device Problem	25/01/2021	25/01/2021
							F12	Serious Injury/ Illness/ Impairment	A24	No Apparent Device Problem	25/01/2021	25/01/2021
							F19	Surgical Intervention	A24	No Apparent Device Problem	25/01/2021	25/01/2021
							F1905	Device Revision or Replacement	A24	No Apparent Device Problem	25/01/2021	25/01/2021
					CT2727	Active implant devices/ancillaries	E2330	Pain	A24	No Apparent Device Problem	25/01/2021	25/01/2021
							F08	Hospitalization or Prolonged Hospitalization	A24	No Apparent Device Problem	25/01/2021	25/01/2021
							F12	Serious Injury/ Illness/ Impairment	A24	No Apparent Device Problem	25/01/2021	25/01/2021
							F19	Surgical Intervention	A24	No Apparent Device Problem	25/01/2021	25/01/2021
							F1905	Device Revision or Replacement	A24	No Apparent Device Problem	25/01/2021	25/01/2021
					CT2750	Implantable cardiac electrical stimulators and associated devices	E2330	Pain	A24	No Apparent Device Problem	25/01/2021	25/01/2021
							F08	Hospitalization or Prolonged Hospitalization	A24	No Apparent Device Problem	25/01/2021	25/01/2021
							F12	Serious Injury/ Illness/ Impairment	A24	No Apparent Device Problem	25/01/2021	25/01/2021
							F19	Surgical Intervention	A24	No Apparent Device Problem	25/01/2021	25/01/2021
							F1905	Device Revision or Replacement	A24	No Apparent Device Problem	25/01/2021	25/01/2021
					CT2866	Implantable cardiac electrical stimulators	E2330	Pain	A24	No Apparent Device Problem	25/01/2021	25/01/2021
							F08	Hospitalization or Prolonged Hospitalization	A24	No Apparent Device Problem	25/01/2021	25/01/2021
							F12	Serious Injury/ Illness/ Impairment	A24	No Apparent Device Problem	25/01/2021	25/01/2021
							F19	Surgical Intervention	A24	No Apparent Device Problem	25/01/2021	25/01/2021
							F1905	Device Revision or Replacement	A24	No Apparent Device Problem	25/01/2021	25/01/2021
					CT2985	Active	E2330	Pain	A24	No Apparent Device Problem	25/01/2021	25/01/2021
							F08	Hospitalization or Prolonged Hospitalization	A24	No Apparent Device Problem	25/01/2021	25/01/2021
							F12	Serious Injury/ Illness/ Impairment	A24	No Apparent Device Problem	25/01/2021	25/01/2021
							F19	Surgical Intervention	A24	No Apparent Device Problem	25/01/2021	25/01/2021
							F1905	Device Revision or Replacement	A24	No Apparent Device Problem	25/01/2021	25/01/2021
					CT3001	Implantable metal	E2330	Pain	A24	No Apparent Device Problem	25/01/2021	25/01/2021
							F08	Hospitalization or Prolonged Hospitalization	A24	No Apparent Device Problem	25/01/2021	25/01/2021
							F12	Serious Injury/ Illness/ Impairment	A24	No Apparent Device Problem	25/01/2021	25/01/2021
							F19	Surgical Intervention	A24	No Apparent Device Problem	25/01/2021	25/01/2021
							F1905	Device Revision or Replacement	A24	No Apparent Device Problem	25/01/2021	25/01/2021
					CT3002	Implantable synthetic polymer	E2330	Pain	A24	No Apparent Device Problem	25/01/2021	25/01/2021
							F08	Hospitalization or Prolonged Hospitalization	A24	No Apparent Device Problem	25/01/2021	25/01/2021
							F12	Serious Injury/ Illness/ Impairment	A24	No Apparent Device Problem	25/01/2021	25/01/2021
							F19	Surgical Intervention	A24	No Apparent Device Problem	25/01/2021	25/01/2021
							F1905	Device Revision or Replacement	A24	No Apparent Device Problem	25/01/2021	25/01/2021
					CT3006	Implantable inorganic material	E2330	Pain	A24	No Apparent Device Problem	25/01/2021	25/01/2021
							F08	Hospitalization or Prolonged Hospitalization	A24	No Apparent Device Problem	25/01/2021	25/01/2021
							F12	Serious Injury/ Illness/ Impairment	A24	No Apparent Device Problem	25/01/2021	25/01/2021
							F19	Surgical Intervention	A24	No Apparent Device Problem	25/01/2021	25/01/2021
							F1905	Device Revision or Replacement	A24	No Apparent Device Problem	25/01/2021	25/01/2021
					CT314	Electrically powered	E2330	Pain	A24	No Apparent Device Problem	25/01/2021	25/01/2021
							F08	Hospitalization or Prolonged Hospitalization	A24	No Apparent Device Problem	25/01/2021	25/01/2021
							F12	Serious Injury/ Illness/ Impairment	A24	No Apparent Device Problem	25/01/2021	25/01/2021
							F19	Surgical Intervention	A24	No Apparent Device Problem	25/01/2021	25/01/2021
							F1905	Device Revision or Replacement	A24	No Apparent Device Problem	25/01/2021	25/01/2021
					CT321	Long-term surgical invasive	E2330	Pain	A24	No Apparent Device Problem	25/01/2021	25/01/2021
							F08	Hospitalization or Prolonged Hospitalization	A24	No Apparent Device Problem	25/01/2021	25/01/2021

							F12	Serious Injury/ Illness/ Impairment	A24	No Apparent Device Problem	25/01/2021	25/01/2021
							F19	Surgical Intervention	A24	No Apparent Device Problem	25/01/2021	25/01/2021
							F1905	Device Revision or Replacement	A24	No Apparent Device Problem	25/01/2021	25/01/2021
					CT449	Generators and associated devices	E2330	Pain	A24	No Apparent Device Problem	25/01/2021	25/01/2021
							F08	Hospitalization or Prolonged Hospitalization	A24	No Apparent Device Problem	25/01/2021	25/01/2021
							F12	Serious Injury/ Illness/ Impairment	A24	No Apparent Device Problem	25/01/2021	25/01/2021
							F19	Surgical Intervention	A24	No Apparent Device Problem	25/01/2021	25/01/2021
							F1905	Device Revision or Replacement	A24	No Apparent Device Problem	25/01/2021	25/01/2021
					CT981	Single use	E2330	Pain	A24	No Apparent Device Problem	25/01/2021	25/01/2021
							F08	Hospitalization or Prolonged Hospitalization	A24	No Apparent Device Problem	25/01/2021	25/01/2021
							F12	Serious Injury/ Illness/ Impairment	A24	No Apparent Device Problem	25/01/2021	25/01/2021
							F19	Surgical Intervention	A24	No Apparent Device Problem	25/01/2021	25/01/2021
							F1905	Device Revision or Replacement	A24	No Apparent Device Problem	25/01/2021	25/01/2021
					CT983	Surgical invasive	E2330	Pain	A24	No Apparent Device Problem	25/01/2021	25/01/2021
							F08	Hospitalization or Prolonged Hospitalization	A24	No Apparent Device Problem	25/01/2021	25/01/2021
							F12	Serious Injury/ Illness/ Impairment	A24	No Apparent Device Problem	25/01/2021	25/01/2021
							F19	Surgical Intervention	A24	No Apparent Device Problem	25/01/2021	25/01/2021
							F1905	Device Revision or Replacement	A24	No Apparent Device Problem	25/01/2021	25/01/2021
10909	2021	All other reportable incidents	36241	Pacemaker/defibrillator lead	CT1684	Cardiac pacemakers and associated devices	E2330	Pain	A24	No Apparent Device Problem	25/01/2021	25/01/2021
							F08	Hospitalization or Prolonged Hospitalization	A24	No Apparent Device Problem	25/01/2021	25/01/2021
							F12	Serious Injury/ Illness/ Impairment	A24	No Apparent Device Problem	25/01/2021	25/01/2021
							F19	Surgical Intervention	A24	No Apparent Device Problem	25/01/2021	25/01/2021
							F1905	Device Revision or Replacement	A24	No Apparent Device Problem	25/01/2021	25/01/2021
10910	2021	All other reportable incidents	36241	Pacemaker/defibrillator lead	CT1684	Cardiac pacemakers and associated devices	E2330	Pain	A24	No Apparent Device Problem	25/01/2021	25/01/2021
							F08	Hospitalization or Prolonged Hospitalization	A24	No Apparent Device Problem	25/01/2021	25/01/2021
							F12	Serious Injury/ Illness/ Impairment	A24	No Apparent Device Problem	25/01/2021	25/01/2021
							F19	Surgical Intervention	A24	No Apparent Device Problem	25/01/2021	25/01/2021
							F1905	Device Revision or Replacement	A24	No Apparent Device Problem	25/01/2021	25/01/2021
10911	2021	All other reportable incidents	47270	Cardiac resynchronization therapy implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A23	Use of Device Problem		
							F08	Hospitalization or Prolonged Hospitalization	A23	Use of Device Problem		
							F1901	Additional Surgery	A23	Use of Device Problem		
10912	2021	All other reportable incidents	36241	Pacemaker/defibrillator lead	CT1684	Cardiac pacemakers and associated devices	E2103	Device Overstimulation of Tissue	A051201	Device Dislodged or Dislocated		
									A070909	Oversensing		
									A0712	Pacing Problem		
							E2403	No Clinical Signs, Symptoms or Conditions	A051201	Device Dislodged or Dislocated		
									A070909	Oversensing		
									A0712	Pacing Problem		
							F26	No Health Consequences or Impact	A051201	Device Dislodged or Dislocated		
									A070909	Oversensing		
									A0712	Pacing Problem		
10913	2021	All other reportable incidents	35852	Single-chamber implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem	24/05/2021	24/05/2021
									A070504	Premature Discharge of Battery	24/05/2021	24/05/2021
							F19	Surgical Intervention	A0705	Battery Problem	24/05/2021	24/05/2021
									A070504	Premature Discharge of Battery	24/05/2021	24/05/2021
10914	2021	All other reportable incidents	35852	Single-chamber implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem	28/04/2021	28/04/2021
									A070504	Premature Discharge of Battery	28/04/2021	28/04/2021
							F19	Surgical Intervention	A0705	Battery Problem	28/04/2021	28/04/2021
									A070504	Premature Discharge of Battery	28/04/2021	28/04/2021
10915	2021	All other reportable incidents	35223	Endocardial/interventricular septal pacing lead	CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A0722	Impedance Problem	07/04/2021	07/04/2021
									A072201	High Impedance	07/04/2021	07/04/2021
									A090801	Signal Artifact/Noise	07/04/2021	07/04/2021
							F1905	Device Revision or Replacement	A0722	Impedance Problem	07/04/2021	07/04/2021
									A072201	High Impedance	07/04/2021	07/04/2021
									A090801	Signal Artifact/Noise	07/04/2021	07/04/2021
							F26	No Health Consequences or Impact	A0722	Impedance Problem	07/04/2021	07/04/2021
									A072201	High Impedance	07/04/2021	07/04/2021
									A090801	Signal Artifact/Noise	07/04/2021	07/04/2021
10916	2021	All other reportable incidents	35852	Single-chamber implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E0133	Stroke/CVA	A070909	Oversensing		
									A071302	Inappropriate/Inadequate Shock/Stimulation		
							E020201	Anxiety	A070909	Oversensing		
									A071302	Inappropriate/Inadequate Shock/Stimulation		
							E2104	Electric Shock	A070909	Oversensing		
									A071302	Inappropriate/Inadequate Shock/Stimulation		
							F08	Hospitalization or Prolonged Hospitalization	A070909	Oversensing		
									A071302	Inappropriate/Inadequate Shock/Stimulation		
							F14	Prolonged Episode of Care	A070909	Oversensing		
									A071302	Inappropriate/Inadequate Shock/Stimulation		
							F23	Unexpected Medical Intervention	A070909	Oversensing		
									A071302	Inappropriate/Inadequate Shock/Stimulation		
10917	2021	All other reportable incidents	17882	Physiologic-monitoring defibrillation system	CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A07	Electrical /Electronic Property Problem		
							F27	Problem Identified During Non-Clinical Procedure	A07	Electrical /Electronic Property Problem		
10918	2021	All other reportable incidents	36241	Pacemaker/defibrillator lead	CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A051201	Device Dislodged or Dislocated		
									A070909	Oversensing		
									A0712	Pacing Problem		
							F11	Non-Serious Injury/ Illness/ Impairment	A051201	Device Dislodged or Dislocated		
									A070909	Oversensing		
									A0712	Pacing Problem		
10919	2021	All other reportable incidents	35852	Single-chamber implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem	14/04/2021	14/04/2021
									A070504	Premature Discharge of Battery	14/04/2021	14/04/2021
							F08	Hospitalization or Prolonged Hospitalization	A0705	Battery Problem	14/04/2021	14/04/2021
									A070504	Premature Discharge of Battery	14/04/2021	14/04/2021
							F1905	Device Revision or Replacement	A0705	Battery Problem	14/04/2021	14/04/2021
									A070504	Premature Discharge of Battery	14/04/2021	14/04/2021
10920	2021	All other reportable incidents	35852	Single-chamber implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	08/04/2021	08/04/2021
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	08/04/2021	08/04/2021
							F19	Surgical Intervention	A070504	Premature Discharge of Battery	08/04/2021	08/04/2021
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	08/04/2021	08/04/2021
10921	2021	All other reportable incidents	36241	Pacemaker/defibrillator lead	CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070909	Oversensing		

							F26	No Health Consequences or Impact	A070909	Oversensing			
10922	2021	All other reportable incidents	36241	Pacemaker/defibrillator lead	CT1684	Cardiac pacemakers and associated devices	E2401	Insufficient Information	A070909	Oversensing			
							E2403	No Clinical Signs, Symptoms or Conditions	A070909	Oversensing			
							F26	No Health Consequences or Impact	A070909	Oversensing			
10923	2021	All other reportable incidents	35852	Single-chamber implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A070909	Oversensing			
							F26	No Health Consequences or Impact	A070909	Oversensing			
10924	2021	All other reportable incidents	47265	Dual-chamber implantable pacemaker, rate-responsive	CT141	Electrodes/Leads and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	01/04/2021	01/04/2021	
							F19	Surgical Intervention	A070504	Premature Discharge of Battery	01/04/2021	01/04/2021	
					CT145	Cardiovascular devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	01/04/2021	01/04/2021	
							F19	Surgical Intervention	A070504	Premature Discharge of Battery	01/04/2021	01/04/2021	
					CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	01/04/2021	01/04/2021	
							F19	Surgical Intervention	A070504	Premature Discharge of Battery	01/04/2021	01/04/2021	
					CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	01/04/2021	01/04/2021	
							F19	Surgical Intervention	A070504	Premature Discharge of Battery	01/04/2021	01/04/2021	
					CT233	Surgical	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	01/04/2021	01/04/2021	
							F19	Surgical Intervention	A070504	Premature Discharge of Battery	01/04/2021	01/04/2021	
					CT2406	Implantable	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	01/04/2021	01/04/2021	
							F19	Surgical Intervention	A070504	Premature Discharge of Battery	01/04/2021	01/04/2021	
					CT2407	Active implantable	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	01/04/2021	01/04/2021	
							F19	Surgical Intervention	A070504	Premature Discharge of Battery	01/04/2021	01/04/2021	
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	01/04/2021	01/04/2021	
							F19	Surgical Intervention	A070504	Premature Discharge of Battery	01/04/2021	01/04/2021	
					CT2727	Active implant devices/ancillaries	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	01/04/2021	01/04/2021	
							F19	Surgical Intervention	A070504	Premature Discharge of Battery	01/04/2021	01/04/2021	
					CT2750	Implantable cardiac electrical stimulators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	01/04/2021	01/04/2021	
							F19	Surgical Intervention	A070504	Premature Discharge of Battery	01/04/2021	01/04/2021	
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	01/04/2021	01/04/2021	
							F19	Surgical Intervention	A070504	Premature Discharge of Battery	01/04/2021	01/04/2021	
					CT2985	Active	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	01/04/2021	01/04/2021	
							F19	Surgical Intervention	A070504	Premature Discharge of Battery	01/04/2021	01/04/2021	
					CT3001	Implantable metal	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	01/04/2021	01/04/2021	
							F19	Surgical Intervention	A070504	Premature Discharge of Battery	01/04/2021	01/04/2021	
					CT3002	Implantable synthetic polymer	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	01/04/2021	01/04/2021	
							F19	Surgical Intervention	A070504	Premature Discharge of Battery	01/04/2021	01/04/2021	
					CT3006	Implantable inorganic material	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	01/04/2021	01/04/2021	
							F19	Surgical Intervention	A070504	Premature Discharge of Battery	01/04/2021	01/04/2021	
					CT314	Electrically powered	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	01/04/2021	01/04/2021	
							F19	Surgical Intervention	A070504	Premature Discharge of Battery	01/04/2021	01/04/2021	
					CT321	Long-term surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	01/04/2021	01/04/2021	
							F19	Surgical Intervention	A070504	Premature Discharge of Battery	01/04/2021	01/04/2021	
					CT449	Generators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	01/04/2021	01/04/2021	
							F19	Surgical Intervention	A070504	Premature Discharge of Battery	01/04/2021	01/04/2021	
					CT981	Single use	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	01/04/2021	01/04/2021	
							F19	Surgical Intervention	A070504	Premature Discharge of Battery	01/04/2021	01/04/2021	
					CT983	Surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	01/04/2021	01/04/2021	
							F19	Surgical Intervention	A070504	Premature Discharge of Battery	01/04/2021	01/04/2021	
10925	2021	All other reportable incidents	47270	Cardiac resynchronization therapy implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	14/04/2021	14/04/2021	
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	14/04/2021	14/04/2021	
							F1903	Device Explantation	A070504	Premature Discharge of Battery	14/04/2021	14/04/2021	
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	14/04/2021	14/04/2021	
10926	2021	All other reportable incidents	47265	Dual-chamber implantable pacemaker, rate-responsive	CT141	Electrodes/Leads and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem	01/07/2021	01/07/2021	
									A070504	Premature Discharge of Battery	01/07/2021	01/07/2021	
							F08	Hospitalization or Prolonged Hospitalization	A0705	Battery Problem	01/07/2021	01/07/2021	
									A070504	Premature Discharge of Battery	01/07/2021	01/07/2021	
							F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A0705	Battery Problem	01/07/2021	01/07/2021	
									A070504	Premature Discharge of Battery	01/07/2021	01/07/2021	
							F1903	Device Explantation	A0705	Battery Problem	01/07/2021	01/07/2021	
									A070504	Premature Discharge of Battery	01/07/2021	01/07/2021	
					CT145	Cardiovascular devices	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem	01/07/2021	01/07/2021	
									A070504	Premature Discharge of Battery	01/07/2021	01/07/2021	
							F08	Hospitalization or Prolonged Hospitalization	A0705	Battery Problem	01/07/2021	01/07/2021	
									A070504	Premature Discharge of Battery	01/07/2021	01/07/2021	
							F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A0705	Battery Problem	01/07/2021	01/07/2021	
									A070504	Premature Discharge of Battery	01/07/2021	01/07/2021	
							F1903	Device Explantation	A0705	Battery Problem	01/07/2021	01/07/2021	
									A070504	Premature Discharge of Battery	01/07/2021	01/07/2021	
					CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem	01/07/2021	01/07/2021	
									A070504	Premature Discharge of Battery	01/07/2021	01/07/2021	
							F08	Hospitalization or Prolonged Hospitalization	A0705	Battery Problem	01/07/2021	01/07/2021	
									A070504	Premature Discharge of Battery	01/07/2021	01/07/2021	
							F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A0705	Battery Problem	01/07/2021	01/07/2021	
									A070504	Premature Discharge of Battery	01/07/2021	01/07/2021	
							F1903	Device Explantation	A0705	Battery Problem	01/07/2021	01/07/2021	
									A070504	Premature Discharge of Battery	01/07/2021	01/07/2021	
					CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem	01/07/2021	01/07/2021	
									A070504	Premature Discharge of Battery	01/07/2021	01/07/2021	
							F08	Hospitalization or Prolonged Hospitalization	A0705	Battery Problem	01/07/2021	01/07/2021	
									A070504	Premature Discharge of Battery	01/07/2021	01/07/2021	
							F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A0705	Battery Problem	01/07/2021	01/07/2021	
									A070504	Premature Discharge of Battery	01/07/2021	01/07/2021	
							F1903	Device Explantation	A0705	Battery Problem	01/07/2021	01/07/2021	
									A070504	Premature Discharge of Battery	01/07/2021	01/07/2021	
					CT233	Surgical	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem	01/07/2021	01/07/2021	
									A070504	Premature Discharge of Battery	01/07/2021	01/07/2021	
							F08	Hospitalization or Prolonged Hospitalization	A0705	Battery Problem	01/07/2021	01/07/2021	
									A070504	Premature Discharge of Battery	01/07/2021	01/07/2021	
							F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A0705	Battery Problem	01/07/2021	01/07/2021	
									A070504	Premature Discharge of Battery	01/07/2021	01/07/2021	

							F1903	Device Explantation	A0705	Battery Problem	01/07/2021	01/07/2021
									A070504	Premature Discharge of Battery	01/07/2021	01/07/2021
					CT2406	Implantable	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem	01/07/2021	01/07/2021
									A070504	Premature Discharge of Battery	01/07/2021	01/07/2021
							F08	Hospitalization or Prolonged Hospitalization	A0705	Battery Problem	01/07/2021	01/07/2021
									A070504	Premature Discharge of Battery	01/07/2021	01/07/2021
							F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A0705	Battery Problem	01/07/2021	01/07/2021
									A070504	Premature Discharge of Battery	01/07/2021	01/07/2021
							F1903	Device Explantation	A0705	Battery Problem	01/07/2021	01/07/2021
									A070504	Premature Discharge of Battery	01/07/2021	01/07/2021
					CT2407	Active implantable	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem	01/07/2021	01/07/2021
									A070504	Premature Discharge of Battery	01/07/2021	01/07/2021
							F08	Hospitalization or Prolonged Hospitalization	A0705	Battery Problem	01/07/2021	01/07/2021
									A070504	Premature Discharge of Battery	01/07/2021	01/07/2021
							F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A0705	Battery Problem	01/07/2021	01/07/2021
									A070504	Premature Discharge of Battery	01/07/2021	01/07/2021
							F1903	Device Explantation	A0705	Battery Problem	01/07/2021	01/07/2021
									A070504	Premature Discharge of Battery	01/07/2021	01/07/2021
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem	01/07/2021	01/07/2021
									A070504	Premature Discharge of Battery	01/07/2021	01/07/2021
							F08	Hospitalization or Prolonged Hospitalization	A0705	Battery Problem	01/07/2021	01/07/2021
									A070504	Premature Discharge of Battery	01/07/2021	01/07/2021
							F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A0705	Battery Problem	01/07/2021	01/07/2021
									A070504	Premature Discharge of Battery	01/07/2021	01/07/2021
							F1903	Device Explantation	A0705	Battery Problem	01/07/2021	01/07/2021
									A070504	Premature Discharge of Battery	01/07/2021	01/07/2021
					CT2727	Active implant devices/ancillaries	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem	01/07/2021	01/07/2021
									A070504	Premature Discharge of Battery	01/07/2021	01/07/2021
							F08	Hospitalization or Prolonged Hospitalization	A0705	Battery Problem	01/07/2021	01/07/2021
									A070504	Premature Discharge of Battery	01/07/2021	01/07/2021
							F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A0705	Battery Problem	01/07/2021	01/07/2021
									A070504	Premature Discharge of Battery	01/07/2021	01/07/2021
							F1903	Device Explantation	A0705	Battery Problem	01/07/2021	01/07/2021
									A070504	Premature Discharge of Battery	01/07/2021	01/07/2021
					CT2750	Implantable cardiac electrical stimulators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem	01/07/2021	01/07/2021
									A070504	Premature Discharge of Battery	01/07/2021	01/07/2021
							F08	Hospitalization or Prolonged Hospitalization	A0705	Battery Problem	01/07/2021	01/07/2021
									A070504	Premature Discharge of Battery	01/07/2021	01/07/2021
							F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A0705	Battery Problem	01/07/2021	01/07/2021
									A070504	Premature Discharge of Battery	01/07/2021	01/07/2021
							F1903	Device Explantation	A0705	Battery Problem	01/07/2021	01/07/2021
									A070504	Premature Discharge of Battery	01/07/2021	01/07/2021
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem	01/07/2021	01/07/2021
									A070504	Premature Discharge of Battery	01/07/2021	01/07/2021
							F08	Hospitalization or Prolonged Hospitalization	A0705	Battery Problem	01/07/2021	01/07/2021
									A070504	Premature Discharge of Battery	01/07/2021	01/07/2021
							F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A0705	Battery Problem	01/07/2021	01/07/2021
									A070504	Premature Discharge of Battery	01/07/2021	01/07/2021
							F1903	Device Explantation	A0705	Battery Problem	01/07/2021	01/07/2021
									A070504	Premature Discharge of Battery	01/07/2021	01/07/2021
					CT2985	Active	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem	01/07/2021	01/07/2021
									A070504	Premature Discharge of Battery	01/07/2021	01/07/2021
							F08	Hospitalization or Prolonged Hospitalization	A0705	Battery Problem	01/07/2021	01/07/2021
									A070504	Premature Discharge of Battery	01/07/2021	01/07/2021
							F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A0705	Battery Problem	01/07/2021	01/07/2021
									A070504	Premature Discharge of Battery	01/07/2021	01/07/2021
							F1903	Device Explantation	A0705	Battery Problem	01/07/2021	01/07/2021
									A070504	Premature Discharge of Battery	01/07/2021	01/07/2021
					CT3001	Implantable metal	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem	01/07/2021	01/07/2021
									A070504	Premature Discharge of Battery	01/07/2021	01/07/2021
							F08	Hospitalization or Prolonged Hospitalization	A0705	Battery Problem	01/07/2021	01/07/2021
									A070504	Premature Discharge of Battery	01/07/2021	01/07/2021
							F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A0705	Battery Problem	01/07/2021	01/07/2021
									A070504	Premature Discharge of Battery	01/07/2021	01/07/2021
							F1903	Device Explantation	A0705	Battery Problem	01/07/2021	01/07/2021
									A070504	Premature Discharge of Battery	01/07/2021	01/07/2021
					CT3002	Implantable synthetic polymer	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem	01/07/2021	01/07/2021
									A070504	Premature Discharge of Battery	01/07/2021	01/07/2021
							F08	Hospitalization or Prolonged Hospitalization	A0705	Battery Problem	01/07/2021	01/07/2021
									A070504	Premature Discharge of Battery	01/07/2021	01/07/2021
							F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A0705	Battery Problem	01/07/2021	01/07/2021
									A070504	Premature Discharge of Battery	01/07/2021	01/07/2021
							F1903	Device Explantation	A0705	Battery Problem	01/07/2021	01/07/2021
									A070504	Premature Discharge of Battery	01/07/2021	01/07/2021
					CT3006	Implantable inorganic material	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem	01/07/2021	01/07/2021
									A070504	Premature Discharge of Battery	01/07/2021	01/07/2021
							F08	Hospitalization or Prolonged Hospitalization	A0705	Battery Problem	01/07/2021	01/07/2021
									A070504	Premature Discharge of Battery	01/07/2021	01/07/2021
							F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A0705	Battery Problem	01/07/2021	01/07/2021
									A070504	Premature Discharge of Battery	01/07/2021	01/07/2021
							F1903	Device Explantation	A0705	Battery Problem	01/07/2021	01/07/2021
									A070504	Premature Discharge of Battery	01/07/2021	01/07/2021
					CT314	Electrically powered	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem	01/07/2021	01/07/2021
									A070504	Premature Discharge of Battery	01/07/2021	01/07/2021
							F08	Hospitalization or Prolonged Hospitalization	A0705	Battery Problem	01/07/2021	01/07/2021
									A070504	Premature Discharge of Battery	01/07/2021	01/07/2021
							F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A0705	Battery Problem	01/07/2021	01/07/2021
									A070504	Premature Discharge of Battery	01/07/2021	01/07/2021

							F1903	Device Explantation	A0705	Battery Problem	01/07/2021	01/07/2021
									A070504	Premature Discharge of Battery	01/07/2021	01/07/2021
					CT321	Long-term surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem	01/07/2021	01/07/2021
									A070504	Premature Discharge of Battery	01/07/2021	01/07/2021
							F08	Hospitalization or Prolonged Hospitalization	A0705	Battery Problem	01/07/2021	01/07/2021
									A070504	Premature Discharge of Battery	01/07/2021	01/07/2021
							F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A0705	Battery Problem	01/07/2021	01/07/2021
									A070504	Premature Discharge of Battery	01/07/2021	01/07/2021
							F1903	Device Explantation	A0705	Battery Problem	01/07/2021	01/07/2021
									A070504	Premature Discharge of Battery	01/07/2021	01/07/2021
					CT449	Generators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem	01/07/2021	01/07/2021
									A070504	Premature Discharge of Battery	01/07/2021	01/07/2021
							F08	Hospitalization or Prolonged Hospitalization	A0705	Battery Problem	01/07/2021	01/07/2021
									A070504	Premature Discharge of Battery	01/07/2021	01/07/2021
							F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A0705	Battery Problem	01/07/2021	01/07/2021
									A070504	Premature Discharge of Battery	01/07/2021	01/07/2021
							F1903	Device Explantation	A0705	Battery Problem	01/07/2021	01/07/2021
									A070504	Premature Discharge of Battery	01/07/2021	01/07/2021
					CT981	Single use	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem	01/07/2021	01/07/2021
									A070504	Premature Discharge of Battery	01/07/2021	01/07/2021
							F08	Hospitalization or Prolonged Hospitalization	A0705	Battery Problem	01/07/2021	01/07/2021
									A070504	Premature Discharge of Battery	01/07/2021	01/07/2021
							F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A0705	Battery Problem	01/07/2021	01/07/2021
									A070504	Premature Discharge of Battery	01/07/2021	01/07/2021
							F1903	Device Explantation	A0705	Battery Problem	01/07/2021	01/07/2021
									A070504	Premature Discharge of Battery	01/07/2021	01/07/2021
					CT983	Surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem	01/07/2021	01/07/2021
									A070504	Premature Discharge of Battery	01/07/2021	01/07/2021
							F08	Hospitalization or Prolonged Hospitalization	A0705	Battery Problem	01/07/2021	01/07/2021
									A070504	Premature Discharge of Battery	01/07/2021	01/07/2021
							F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A0705	Battery Problem	01/07/2021	01/07/2021
									A070504	Premature Discharge of Battery	01/07/2021	01/07/2021
							F1903	Device Explantation	A0705	Battery Problem	01/07/2021	01/07/2021
									A070504	Premature Discharge of Battery	01/07/2021	01/07/2021
10927	2021	All other reportable incidents	37265	Dual-chamber implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A160601	Premature Elective Replacement Indicator		
							F1905	Device Revision or Replacement	A160601	Premature Elective Replacement Indicator		
10928	2021	All other reportable incidents	35852	Single-chamber implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E060102	Atrial Fibrillation	A1201	Blocked Connection		
							F1905	Device Revision or Replacement	A1201	Blocked Connection		
10929	2021	All other reportable incidents	35852	Single-chamber implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F1901	Additional Surgery	A070504	Premature Discharge of Battery		
10930	2021	All other reportable incidents	17882	Physiologic-monitoring defibrillation system	CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070802	Intermittent Loss of Power		
							F27	Problem Identified During Non-Clinical Procedure	A070802	Intermittent Loss of Power		
10931	2021	All other reportable incidents	47270	Cardiac resynchronization therapy implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A072201	High Impedance		
							F26	No Health Consequences or Impact	A072201	High Impedance		
10932	2021	All other reportable incidents	36241	Pacemaker/defibrillator lead	CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A072201	High Impedance		
							F26	No Health Consequences or Impact	A072201	High Impedance		
10933	2021	All other reportable incidents	35852	Single-chamber implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E060102	Atrial Fibrillation	A010402	Migration	06/05/2021	06/05/2021
									A071201	Failure to Convert Rhythm	06/05/2021	06/05/2021
									A072201	High Impedance	06/05/2021	06/05/2021
							E060110	Ventricular Fibrillation	A010402	Migration	06/05/2021	06/05/2021
									A071201	Failure to Convert Rhythm	06/05/2021	06/05/2021
									A072201	High Impedance	06/05/2021	06/05/2021
							E0611	Heart Failure/Congestive Heart Failure	A010402	Migration	06/05/2021	06/05/2021
									A071201	Failure to Convert Rhythm	06/05/2021	06/05/2021
									A072201	High Impedance	06/05/2021	06/05/2021
							F05	Delay to Treatment/ Therapy	A010402	Migration	06/05/2021	06/05/2021
									A071201	Failure to Convert Rhythm	06/05/2021	06/05/2021
									A072201	High Impedance	06/05/2021	06/05/2021
							F1001	Absence of Treatment	A010402	Migration	06/05/2021	06/05/2021
									A071201	Failure to Convert Rhythm	06/05/2021	06/05/2021
									A072201	High Impedance	06/05/2021	06/05/2021
							F19	Surgical Intervention	A010402	Migration	06/05/2021	06/05/2021
									A071201	Failure to Convert Rhythm	06/05/2021	06/05/2021
									A072201	High Impedance	06/05/2021	06/05/2021
10934	2021	All other reportable incidents	35852	Single-chamber implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A010402	Migration		
							F08	Hospitalization or Prolonged Hospitalization	A010402	Migration		
							F1905	Device Revision or Replacement	A010402	Migration		
10935	2021	All other reportable incidents	47265	Dual-chamber implantable pacemaker, rate-responsive	CT141	Electrodes/Leads and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem	14/06/2021	14/06/2021
									A070504	Premature Discharge of Battery	14/06/2021	14/06/2021
							F08	Hospitalization or Prolonged Hospitalization	A0705	Battery Problem	14/06/2021	14/06/2021
									A070504	Premature Discharge of Battery	14/06/2021	14/06/2021
							F1905	Device Revision or Replacement	A0705	Battery Problem	14/06/2021	14/06/2021
									A070504	Premature Discharge of Battery	14/06/2021	14/06/2021
					CT145	Cardiovascular devices	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem	14/06/2021	14/06/2021
									A070504	Premature Discharge of Battery	14/06/2021	14/06/2021
							F08	Hospitalization or Prolonged Hospitalization	A0705	Battery Problem	14/06/2021	14/06/2021
									A070504	Premature Discharge of Battery	14/06/2021	14/06/2021
							F1905	Device Revision or Replacement	A0705	Battery Problem	14/06/2021	14/06/2021
									A070504	Premature Discharge of Battery	14/06/2021	14/06/2021
					CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem	14/06/2021	14/06/2021
									A070504	Premature Discharge of Battery	14/06/2021	14/06/2021
							F08	Hospitalization or Prolonged Hospitalization	A0705	Battery Problem	14/06/2021	14/06/2021
									A070504	Premature Discharge of Battery	14/06/2021	14/06/2021
							F1905	Device Revision or Replacement	A0705	Battery Problem	14/06/2021	14/06/2021
									A070504	Premature Discharge of Battery	14/06/2021	14/06/2021
					CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem	14/06/2021	14/06/2021
									A070504	Premature Discharge of Battery	14/06/2021	14/06/2021
							F08	Hospitalization or Prolonged Hospitalization	A0705	Battery Problem	14/06/2021	14/06/2021

								A070504	Premature Discharge of Battery	14/06/2021	14/06/2021
						F1905	Device Revision or Replacement	A0705	Battery Problem	14/06/2021	14/06/2021
								A070504	Premature Discharge of Battery	14/06/2021	14/06/2021
					CT233	Surgical	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem	14/06/2021	14/06/2021
									A070504	Premature Discharge of Battery	14/06/2021	14/06/2021
							F08	Hospitalization or Prolonged Hospitalization	A0705	Battery Problem	14/06/2021	14/06/2021
									A070504	Premature Discharge of Battery	14/06/2021	14/06/2021
							F1905	Device Revision or Replacement	A0705	Battery Problem	14/06/2021	14/06/2021
									A070504	Premature Discharge of Battery	14/06/2021	14/06/2021
					CT2406	Implantable	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem	14/06/2021	14/06/2021
									A070504	Premature Discharge of Battery	14/06/2021	14/06/2021
							F08	Hospitalization or Prolonged Hospitalization	A0705	Battery Problem	14/06/2021	14/06/2021
									A070504	Premature Discharge of Battery	14/06/2021	14/06/2021
							F1905	Device Revision or Replacement	A0705	Battery Problem	14/06/2021	14/06/2021
									A070504	Premature Discharge of Battery	14/06/2021	14/06/2021
					CT2407	Active implantable	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem	14/06/2021	14/06/2021
									A070504	Premature Discharge of Battery	14/06/2021	14/06/2021
							F08	Hospitalization or Prolonged Hospitalization	A0705	Battery Problem	14/06/2021	14/06/2021
									A070504	Premature Discharge of Battery	14/06/2021	14/06/2021
							F1905	Device Revision or Replacement	A0705	Battery Problem	14/06/2021	14/06/2021
									A070504	Premature Discharge of Battery	14/06/2021	14/06/2021
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem	14/06/2021	14/06/2021
									A070504	Premature Discharge of Battery	14/06/2021	14/06/2021
							F08	Hospitalization or Prolonged Hospitalization	A0705	Battery Problem	14/06/2021	14/06/2021
									A070504	Premature Discharge of Battery	14/06/2021	14/06/2021
							F1905	Device Revision or Replacement	A0705	Battery Problem	14/06/2021	14/06/2021
									A070504	Premature Discharge of Battery	14/06/2021	14/06/2021
					CT2727	Active implant devices/ancillaries	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem	14/06/2021	14/06/2021
									A070504	Premature Discharge of Battery	14/06/2021	14/06/2021
							F08	Hospitalization or Prolonged Hospitalization	A0705	Battery Problem	14/06/2021	14/06/2021
									A070504	Premature Discharge of Battery	14/06/2021	14/06/2021
							F1905	Device Revision or Replacement	A0705	Battery Problem	14/06/2021	14/06/2021
									A070504	Premature Discharge of Battery	14/06/2021	14/06/2021
					CT2750	Implantable cardiac electrical stimulators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem	14/06/2021	14/06/2021
									A070504	Premature Discharge of Battery	14/06/2021	14/06/2021
							F08	Hospitalization or Prolonged Hospitalization	A0705	Battery Problem	14/06/2021	14/06/2021
									A070504	Premature Discharge of Battery	14/06/2021	14/06/2021
							F1905	Device Revision or Replacement	A0705	Battery Problem	14/06/2021	14/06/2021
									A070504	Premature Discharge of Battery	14/06/2021	14/06/2021
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem	14/06/2021	14/06/2021
									A070504	Premature Discharge of Battery	14/06/2021	14/06/2021
							F08	Hospitalization or Prolonged Hospitalization	A0705	Battery Problem	14/06/2021	14/06/2021
									A070504	Premature Discharge of Battery	14/06/2021	14/06/2021
							F1905	Device Revision or Replacement	A0705	Battery Problem	14/06/2021	14/06/2021
									A070504	Premature Discharge of Battery	14/06/2021	14/06/2021
					CT2985	Active	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem	14/06/2021	14/06/2021
									A070504	Premature Discharge of Battery	14/06/2021	14/06/2021
							F08	Hospitalization or Prolonged Hospitalization	A0705	Battery Problem	14/06/2021	14/06/2021
									A070504	Premature Discharge of Battery	14/06/2021	14/06/2021
							F1905	Device Revision or Replacement	A0705	Battery Problem	14/06/2021	14/06/2021
									A070504	Premature Discharge of Battery	14/06/2021	14/06/2021
					CT3001	Implantable metal	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem	14/06/2021	14/06/2021
									A070504	Premature Discharge of Battery	14/06/2021	14/06/2021
							F08	Hospitalization or Prolonged Hospitalization	A0705	Battery Problem	14/06/2021	14/06/2021
									A070504	Premature Discharge of Battery	14/06/2021	14/06/2021
							F1905	Device Revision or Replacement	A0705	Battery Problem	14/06/2021	14/06/2021
									A070504	Premature Discharge of Battery	14/06/2021	14/06/2021
					CT3002	Implantable synthetic polymer	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem	14/06/2021	14/06/2021
									A070504	Premature Discharge of Battery	14/06/2021	14/06/2021
							F08	Hospitalization or Prolonged Hospitalization	A0705	Battery Problem	14/06/2021	14/06/2021
									A070504	Premature Discharge of Battery	14/06/2021	14/06/2021
							F1905	Device Revision or Replacement	A0705	Battery Problem	14/06/2021	14/06/2021
									A070504	Premature Discharge of Battery	14/06/2021	14/06/2021
					CT3006	Implantable inorganic material	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem	14/06/2021	14/06/2021
									A070504	Premature Discharge of Battery	14/06/2021	14/06/2021
							F08	Hospitalization or Prolonged Hospitalization	A0705	Battery Problem	14/06/2021	14/06/2021
									A070504	Premature Discharge of Battery	14/06/2021	14/06/2021
							F1905	Device Revision or Replacement	A0705	Battery Problem	14/06/2021	14/06/2021
									A070504	Premature Discharge of Battery	14/06/2021	14/06/2021
					CT314	Electrically powered	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem	14/06/2021	14/06/2021
									A070504	Premature Discharge of Battery	14/06/2021	14/06/2021
							F08	Hospitalization or Prolonged Hospitalization	A0705	Battery Problem	14/06/2021	14/06/2021
									A070504	Premature Discharge of Battery	14/06/2021	14/06/2021
							F1905	Device Revision or Replacement	A0705	Battery Problem	14/06/2021	14/06/2021
									A070504	Premature Discharge of Battery	14/06/2021	14/06/2021
					CT321	Long-term surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem	14/06/2021	14/06/2021
									A070504	Premature Discharge of Battery	14/06/2021	14/06/2021
							F08	Hospitalization or Prolonged Hospitalization	A0705	Battery Problem	14/06/2021	14/06/2021
									A070504	Premature Discharge of Battery	14/06/2021	14/06/2021
							F1905	Device Revision or Replacement	A0705	Battery Problem	14/06/2021	14/06/2021
									A070504	Premature Discharge of Battery	14/06/2021	14/06/2021
					CT449	Generators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem	14/06/2021	14/06/2021
									A070504	Premature Discharge of Battery	14/06/2021	14/06/2021
							F08	Hospitalization or Prolonged Hospitalization	A0705	Battery Problem	14/06/2021	14/06/2021
									A070504	Premature Discharge of Battery	14/06/2021	14/06/2021
							F1905	Device Revision or Replacement	A0705	Battery Problem	14/06/2021	14/06/2021
									A070504	Premature Discharge of Battery	14/06/2021	14/06/2021
					CT981	Single use	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem	14/06/2021	14/06/2021

									A070504	Premature Discharge of Battery	14/06/2021	14/06/2021
							F08	Hospitalization or Prolonged Hospitalization	A0705	Battery Problem	14/06/2021	14/06/2021
									A070504	Premature Discharge of Battery	14/06/2021	14/06/2021
							F1905	Device Revision or Replacement	A0705	Battery Problem	14/06/2021	14/06/2021
									A070504	Premature Discharge of Battery	14/06/2021	14/06/2021
					CT983	Surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem	14/06/2021	14/06/2021
									A070504	Premature Discharge of Battery	14/06/2021	14/06/2021
							F08	Hospitalization or Prolonged Hospitalization	A0705	Battery Problem	14/06/2021	14/06/2021
									A070504	Premature Discharge of Battery	14/06/2021	14/06/2021
							F1905	Device Revision or Replacement	A0705	Battery Problem	14/06/2021	14/06/2021
									A070504	Premature Discharge of Battery	14/06/2021	14/06/2021
10936	2021	All other reportable incidents	35852	Single-chamber implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E2104	Electric Shock	A071302	Inappropriate/Inadequate Shock/Stimulation		
							E2403	No Clinical Signs, Symptoms or Conditions	A071302	Inappropriate/Inadequate Shock/Stimulation		
							F26	No Health Consequences or Impact	A071302	Inappropriate/Inadequate Shock/Stimulation		
10937	2021	All other reportable incidents	47267	Single-chamber implantable pacemaker, rate-responsive	CT141	Electrodes/Leads and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	07/04/2021	07/04/2021
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	07/04/2021	07/04/2021
							F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A070504	Premature Discharge of Battery	07/04/2021	07/04/2021
							F12	Serious Injury/ Illness/ Impairment	A070504	Premature Discharge of Battery	07/04/2021	07/04/2021
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	07/04/2021	07/04/2021
					CT145	Cardiovascular devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	07/04/2021	07/04/2021
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	07/04/2021	07/04/2021
							F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A070504	Premature Discharge of Battery	07/04/2021	07/04/2021
							F12	Serious Injury/ Illness/ Impairment	A070504	Premature Discharge of Battery	07/04/2021	07/04/2021
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	07/04/2021	07/04/2021
					CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	07/04/2021	07/04/2021
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	07/04/2021	07/04/2021
							F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A070504	Premature Discharge of Battery	07/04/2021	07/04/2021
							F12	Serious Injury/ Illness/ Impairment	A070504	Premature Discharge of Battery	07/04/2021	07/04/2021
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	07/04/2021	07/04/2021
					CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	07/04/2021	07/04/2021
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	07/04/2021	07/04/2021
							F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A070504	Premature Discharge of Battery	07/04/2021	07/04/2021
							F12	Serious Injury/ Illness/ Impairment	A070504	Premature Discharge of Battery	07/04/2021	07/04/2021
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	07/04/2021	07/04/2021
					CT233	Surgical	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	07/04/2021	07/04/2021
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	07/04/2021	07/04/2021
							F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A070504	Premature Discharge of Battery	07/04/2021	07/04/2021
							F12	Serious Injury/ Illness/ Impairment	A070504	Premature Discharge of Battery	07/04/2021	07/04/2021
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	07/04/2021	07/04/2021
					CT2406	Implantable	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	07/04/2021	07/04/2021
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	07/04/2021	07/04/2021
							F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A070504	Premature Discharge of Battery	07/04/2021	07/04/2021
							F12	Serious Injury/ Illness/ Impairment	A070504	Premature Discharge of Battery	07/04/2021	07/04/2021
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	07/04/2021	07/04/2021
					CT2407	Active implantable	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	07/04/2021	07/04/2021
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	07/04/2021	07/04/2021
							F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A070504	Premature Discharge of Battery	07/04/2021	07/04/2021
							F12	Serious Injury/ Illness/ Impairment	A070504	Premature Discharge of Battery	07/04/2021	07/04/2021
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	07/04/2021	07/04/2021
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	07/04/2021	07/04/2021
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	07/04/2021	07/04/2021
							F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A070504	Premature Discharge of Battery	07/04/2021	07/04/2021
							F12	Serious Injury/ Illness/ Impairment	A070504	Premature Discharge of Battery	07/04/2021	07/04/2021
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	07/04/2021	07/04/2021
					CT2727	Active implant devices/ancillaries	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	07/04/2021	07/04/2021
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	07/04/2021	07/04/2021
							F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A070504	Premature Discharge of Battery	07/04/2021	07/04/2021
							F12	Serious Injury/ Illness/ Impairment	A070504	Premature Discharge of Battery	07/04/2021	07/04/2021
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	07/04/2021	07/04/2021
					CT2750	Implantable cardiac electrical stimulators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	07/04/2021	07/04/2021
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	07/04/2021	07/04/2021
							F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A070504	Premature Discharge of Battery	07/04/2021	07/04/2021
							F12	Serious Injury/ Illness/ Impairment	A070504	Premature Discharge of Battery	07/04/2021	07/04/2021
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	07/04/2021	07/04/2021
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	07/04/2021	07/04/2021
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	07/04/2021	07/04/2021
							F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A070504	Premature Discharge of Battery	07/04/2021	07/04/2021
							F12	Serious Injury/ Illness/ Impairment	A070504	Premature Discharge of Battery	07/04/2021	07/04/2021
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	07/04/2021	07/04/2021
					CT2985	Active	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	07/04/2021	07/04/2021
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	07/04/2021	07/04/2021
							F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A070504	Premature Discharge of Battery	07/04/2021	07/04/2021
							F12	Serious Injury/ Illness/ Impairment	A070504	Premature Discharge of Battery	07/04/2021	07/04/2021
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	07/04/2021	07/04/2021
					CT3001	Implantable metal	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	07/04/2021	07/04/2021
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	07/04/2021	07/04/2021
							F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A070504	Premature Discharge of Battery	07/04/2021	07/04/2021
							F12	Serious Injury/ Illness/ Impairment	A070504	Premature Discharge of Battery	07/04/2021	07/04/2021
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	07/04/2021	07/04/2021
					CT3002	Implantable synthetic polymer	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	07/04/2021	07/04/2021
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	07/04/2021	07/04/2021
							F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A070504	Premature Discharge of Battery	07/04/2021	07/04/2021
							F12	Serious Injury/ Illness/ Impairment	A070504	Premature Discharge of Battery	07/04/2021	07/04/2021
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	07/04/2021	07/04/2021
					CT3006	Implantable inorganic material	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	07/04/2021	07/04/2021
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	07/04/2021	07/04/2021
							F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A070504	Premature Discharge of Battery	07/04/2021	07/04/2021
							F12	Serious Injury/ Illness/ Impairment	A070504	Premature Discharge of Battery	07/04/2021	07/04/2021

							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	07/04/2021	07/04/2021
					CT314	Electrically powered	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	07/04/2021	07/04/2021
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	07/04/2021	07/04/2021
							F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A070504	Premature Discharge of Battery	07/04/2021	07/04/2021
							F12	Serious Injury/ Illness/ Impairment	A070504	Premature Discharge of Battery	07/04/2021	07/04/2021
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	07/04/2021	07/04/2021
					CT321	Long-term surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	07/04/2021	07/04/2021
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	07/04/2021	07/04/2021
							F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A070504	Premature Discharge of Battery	07/04/2021	07/04/2021
							F12	Serious Injury/ Illness/ Impairment	A070504	Premature Discharge of Battery	07/04/2021	07/04/2021
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	07/04/2021	07/04/2021
					CT449	Generators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	07/04/2021	07/04/2021
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	07/04/2021	07/04/2021
							F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A070504	Premature Discharge of Battery	07/04/2021	07/04/2021
							F12	Serious Injury/ Illness/ Impairment	A070504	Premature Discharge of Battery	07/04/2021	07/04/2021
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	07/04/2021	07/04/2021
					CT981	Single use	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	07/04/2021	07/04/2021
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	07/04/2021	07/04/2021
							F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A070504	Premature Discharge of Battery	07/04/2021	07/04/2021
							F12	Serious Injury/ Illness/ Impairment	A070504	Premature Discharge of Battery	07/04/2021	07/04/2021
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	07/04/2021	07/04/2021
					CT983	Surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	07/04/2021	07/04/2021
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	07/04/2021	07/04/2021
							F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A070504	Premature Discharge of Battery	07/04/2021	07/04/2021
							F12	Serious Injury/ Illness/ Impairment	A070504	Premature Discharge of Battery	07/04/2021	07/04/2021
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	07/04/2021	07/04/2021
10938	2021	All other reportable incidents	37265	Dual-chamber implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E2401	Insufficient Information	A070504	Premature Discharge of Battery		
							E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
10939	2021	All other reportable incidents	47263	Cardiac resynchronization therapy implantable pacemaker	CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A0712	Pacing Problem		
							F26	No Health Consequences or Impact	A0712	Pacing Problem		
					CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A0712	Pacing Problem		
							F26	No Health Consequences or Impact	A0712	Pacing Problem		
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A0712	Pacing Problem		
							F26	No Health Consequences or Impact	A0712	Pacing Problem		
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A0712	Pacing Problem		
							F26	No Health Consequences or Impact	A0712	Pacing Problem		
10940	2021	All other reportable incidents	35852	Single-chamber implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A072201	High Impedance		
							F26	No Health Consequences or Impact	A072201	High Impedance		
10941	2021	All other reportable incidents	35852	Single-chamber implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E2104	Electric Shock	A070909	Oversensing		
									A071302	Inappropriate/Inadequate Shock/Stimulation		
							E2403	No Clinical Signs, Symptoms or Conditions	A070909	Oversensing		
									A071302	Inappropriate/Inadequate Shock/Stimulation		
							F26	No Health Consequences or Impact	A070909	Oversensing		
									A071302	Inappropriate/Inadequate Shock/Stimulation		
10942	2021	All other reportable incidents	37265	Dual-chamber implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E2311	Discomfort	A071302	Inappropriate/Inadequate Shock/Stimulation		
							E2330	Pain	A071302	Inappropriate/Inadequate Shock/Stimulation		
							F23	Unexpected Medical Intervention	A071302	Inappropriate/Inadequate Shock/Stimulation		
10943	2021	All other reportable incidents	17882	Physiologic-monitoring defibrillation system	CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A07	Electrical /Electronic Property Problem		
							F27	Problem Identified During Non-Clinical Procedure	A07	Electrical /Electronic Property Problem		
10944	2021	All other reportable incidents	36241	Pacemaker/defibrillator lead	CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A040101	Fracture		
									A072201	High Impedance		
							F08	Hospitalization or Prolonged Hospitalization	A040101	Fracture		
									A072201	High Impedance		
							F1905	Device Revision or Replacement	A040101	Fracture		
									A072201	High Impedance		
10945	2021	All other reportable incidents	17882	Physiologic-monitoring defibrillation system	CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A0719	Unexpected Shutdown		
							F26	No Health Consequences or Impact	A0719	Unexpected Shutdown		
10946	2021	All other reportable incidents	60910	Coronary venous pacing lead	CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A0722	Impedance Problem		
							F26	No Health Consequences or Impact	A0722	Impedance Problem		
10947	2021	All other reportable incidents	47265	Dual-chamber implantable pacemaker, rate-responsive	CT141	Electrodes/Leads and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT145	Cardiovascular devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT233	Surgical	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT2406	Implantable	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT2407	Active implantable	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT2727	Active implant devices/ancillaries	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT2750	Implantable cardiac electrical stimulators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		

							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT2985	Active	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT3001	Implantable metal	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT3002	Implantable synthetic polymer	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT3006	Implantable inorganic material	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT314	Electrically powered	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT321	Long-term surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT449	Generators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT981	Single use	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT983	Surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
10948	2021	All other reportable incidents	47263	Cardiac resynchronization therapy implantable pacemaker	CT1684	Cardiac pacemakers and associated devices	E011903	Syncope/Fainting	A0712	Pacing Problem		
							E2403	No Clinical Signs, Symptoms or Conditions	A0712	Pacing Problem		
							F08	Hospitalization or Prolonged Hospitalization	A0712	Pacing Problem		
							F15	Recognized Device or Procedural Complication	A0712	Pacing Problem		
					CT1685	Cardiac pacemakers	E011903	Syncope/Fainting	A0712	Pacing Problem		
							E2403	No Clinical Signs, Symptoms or Conditions	A0712	Pacing Problem		
							F08	Hospitalization or Prolonged Hospitalization	A0712	Pacing Problem		
							F15	Recognized Device or Procedural Complication	A0712	Pacing Problem		
					CT257	Implantable cardiac pacemakers	E011903	Syncope/Fainting	A0712	Pacing Problem		
							E2403	No Clinical Signs, Symptoms or Conditions	A0712	Pacing Problem		
							F08	Hospitalization or Prolonged Hospitalization	A0712	Pacing Problem		
							F15	Recognized Device or Procedural Complication	A0712	Pacing Problem		
					CT2866	Implantable cardiac electrical stimulators	E011903	Syncope/Fainting	A0712	Pacing Problem		
							E2403	No Clinical Signs, Symptoms or Conditions	A0712	Pacing Problem		
							F08	Hospitalization or Prolonged Hospitalization	A0712	Pacing Problem		
							F15	Recognized Device or Procedural Complication	A0712	Pacing Problem		
10949	2021	All other reportable incidents	35852	Single-chamber implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	15/04/2021	15/04/2021
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	15/04/2021	15/04/2021
							F12	Serious Injury/ Illness/ Impairment	A070504	Premature Discharge of Battery	15/04/2021	15/04/2021
							F19	Surgical Intervention	A070504	Premature Discharge of Battery	15/04/2021	15/04/2021
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	15/04/2021	15/04/2021
10950	2021	All other reportable incidents	60910	Coronary venous pacing lead	CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A0401	Break		
							F26	No Health Consequences or Impact	A0401	Break		
10951	2021	All other reportable incidents	47270	Cardiac resynchronization therapy implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A090805	Incorrect Measurement	16/04/2021	16/04/2021
							F1901	Additional Surgery	A090805	Incorrect Measurement	16/04/2021	16/04/2021
10952	2021	All other reportable incidents	35852	Single-chamber implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	19/05/2021	19/05/2021
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	19/05/2021	19/05/2021
							F12	Serious Injury/ Illness/ Impairment	A070504	Premature Discharge of Battery	19/05/2021	19/05/2021
							F19	Surgical Intervention	A070504	Premature Discharge of Battery	19/05/2021	19/05/2021
10953	2021	All other reportable incidents	47265	Dual-chamber implantable pacemaker, rate-responsive	CT141	Electrodes/Leads and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A0908	Incorrect, Inadequate or Imprecise Result or Readings	26/04/2021	26/04/2021
							F08	Hospitalization or Prolonged Hospitalization	A0908	Incorrect, Inadequate or Imprecise Result or Readings	26/04/2021	26/04/2021
							F19	Surgical Intervention	A0908	Incorrect, Inadequate or Imprecise Result or Readings	26/04/2021	26/04/2021
							F1903	Device Explantation	A0908	Incorrect, Inadequate or Imprecise Result or Readings	26/04/2021	26/04/2021
					CT145	Cardiovascular devices	E2403	No Clinical Signs, Symptoms or Conditions	A0908	Incorrect, Inadequate or Imprecise Result or Readings	26/04/2021	26/04/2021
							F08	Hospitalization or Prolonged Hospitalization	A0908	Incorrect, Inadequate or Imprecise Result or Readings	26/04/2021	26/04/2021
							F19	Surgical Intervention	A0908	Incorrect, Inadequate or Imprecise Result or Readings	26/04/2021	26/04/2021
							F1903	Device Explantation	A0908	Incorrect, Inadequate or Imprecise Result or Readings	26/04/2021	26/04/2021
					CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A0908	Incorrect, Inadequate or Imprecise Result or Readings	26/04/2021	26/04/2021
							F08	Hospitalization or Prolonged Hospitalization	A0908	Incorrect, Inadequate or Imprecise Result or Readings	26/04/2021	26/04/2021
							F19	Surgical Intervention	A0908	Incorrect, Inadequate or Imprecise Result or Readings	26/04/2021	26/04/2021
							F1903	Device Explantation	A0908	Incorrect, Inadequate or Imprecise Result or Readings	26/04/2021	26/04/2021
					CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A0908	Incorrect, Inadequate or Imprecise Result or Readings	26/04/2021	26/04/2021
							F08	Hospitalization or Prolonged Hospitalization	A0908	Incorrect, Inadequate or Imprecise Result or Readings	26/04/2021	26/04/2021
							F19	Surgical Intervention	A0908	Incorrect, Inadequate or Imprecise Result or Readings	26/04/2021	26/04/2021

							F1903	Device Explantation	A0908	Incorrect, Inadequate or Imprecise Result or Readings	26/04/2021	26/04/2021
					CT233	Surgical	E2403	No Clinical Signs, Symptoms or Conditions	A0908	Incorrect, Inadequate or Imprecise Result or Readings	26/04/2021	26/04/2021
							F08	Hospitalization or Prolonged Hospitalization	A0908	Incorrect, Inadequate or Imprecise Result or Readings	26/04/2021	26/04/2021
							F19	Surgical Intervention	A0908	Incorrect, Inadequate or Imprecise Result or Readings	26/04/2021	26/04/2021
							F1903	Device Explantation	A0908	Incorrect, Inadequate or Imprecise Result or Readings	26/04/2021	26/04/2021
					CT2406	Implantable	E2403	No Clinical Signs, Symptoms or Conditions	A0908	Incorrect, Inadequate or Imprecise Result or Readings	26/04/2021	26/04/2021
							F08	Hospitalization or Prolonged Hospitalization	A0908	Incorrect, Inadequate or Imprecise Result or Readings	26/04/2021	26/04/2021
							F19	Surgical Intervention	A0908	Incorrect, Inadequate or Imprecise Result or Readings	26/04/2021	26/04/2021
							F1903	Device Explantation	A0908	Incorrect, Inadequate or Imprecise Result or Readings	26/04/2021	26/04/2021
					CT2407	Active implantable	E2403	No Clinical Signs, Symptoms or Conditions	A0908	Incorrect, Inadequate or Imprecise Result or Readings	26/04/2021	26/04/2021
							F08	Hospitalization or Prolonged Hospitalization	A0908	Incorrect, Inadequate or Imprecise Result or Readings	26/04/2021	26/04/2021
							F19	Surgical Intervention	A0908	Incorrect, Inadequate or Imprecise Result or Readings	26/04/2021	26/04/2021
							F1903	Device Explantation	A0908	Incorrect, Inadequate or Imprecise Result or Readings	26/04/2021	26/04/2021
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A0908	Incorrect, Inadequate or Imprecise Result or Readings	26/04/2021	26/04/2021
							F08	Hospitalization or Prolonged Hospitalization	A0908	Incorrect, Inadequate or Imprecise Result or Readings	26/04/2021	26/04/2021
							F19	Surgical Intervention	A0908	Incorrect, Inadequate or Imprecise Result or Readings	26/04/2021	26/04/2021
							F1903	Device Explantation	A0908	Incorrect, Inadequate or Imprecise Result or Readings	26/04/2021	26/04/2021
					CT2727	Active implant devices/ancillaries	E2403	No Clinical Signs, Symptoms or Conditions	A0908	Incorrect, Inadequate or Imprecise Result or Readings	26/04/2021	26/04/2021
							F08	Hospitalization or Prolonged Hospitalization	A0908	Incorrect, Inadequate or Imprecise Result or Readings	26/04/2021	26/04/2021
							F19	Surgical Intervention	A0908	Incorrect, Inadequate or Imprecise Result or Readings	26/04/2021	26/04/2021
							F1903	Device Explantation	A0908	Incorrect, Inadequate or Imprecise Result or Readings	26/04/2021	26/04/2021
					CT2750	Implantable cardiac electrical stimulators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A0908	Incorrect, Inadequate or Imprecise Result or Readings	26/04/2021	26/04/2021
							F08	Hospitalization or Prolonged Hospitalization	A0908	Incorrect, Inadequate or Imprecise Result or Readings	26/04/2021	26/04/2021
							F19	Surgical Intervention	A0908	Incorrect, Inadequate or Imprecise Result or Readings	26/04/2021	26/04/2021
							F1903	Device Explantation	A0908	Incorrect, Inadequate or Imprecise Result or Readings	26/04/2021	26/04/2021
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A0908	Incorrect, Inadequate or Imprecise Result or Readings	26/04/2021	26/04/2021
							F08	Hospitalization or Prolonged Hospitalization	A0908	Incorrect, Inadequate or Imprecise Result or Readings	26/04/2021	26/04/2021
							F19	Surgical Intervention	A0908	Incorrect, Inadequate or Imprecise Result or Readings	26/04/2021	26/04/2021
							F1903	Device Explantation	A0908	Incorrect, Inadequate or Imprecise Result or Readings	26/04/2021	26/04/2021
					CT2985	Active	E2403	No Clinical Signs, Symptoms or Conditions	A0908	Incorrect, Inadequate or Imprecise Result or Readings	26/04/2021	26/04/2021
							F08	Hospitalization or Prolonged Hospitalization	A0908	Incorrect, Inadequate or Imprecise Result or Readings	26/04/2021	26/04/2021
							F19	Surgical Intervention	A0908	Incorrect, Inadequate or Imprecise Result or Readings	26/04/2021	26/04/2021
							F1903	Device Explantation	A0908	Incorrect, Inadequate or Imprecise Result or Readings	26/04/2021	26/04/2021
					CT3001	Implantable metal	E2403	No Clinical Signs, Symptoms or Conditions	A0908	Incorrect, Inadequate or Imprecise Result or Readings	26/04/2021	26/04/2021
							F08	Hospitalization or Prolonged Hospitalization	A0908	Incorrect, Inadequate or Imprecise Result or Readings	26/04/2021	26/04/2021
							F19	Surgical Intervention	A0908	Incorrect, Inadequate or Imprecise Result or Readings	26/04/2021	26/04/2021
							F1903	Device Explantation	A0908	Incorrect, Inadequate or Imprecise Result or Readings	26/04/2021	26/04/2021
					CT3002	Implantable synthetic polymer	E2403	No Clinical Signs, Symptoms or Conditions	A0908	Incorrect, Inadequate or Imprecise Result or Readings	26/04/2021	26/04/2021
							F08	Hospitalization or Prolonged Hospitalization	A0908	Incorrect, Inadequate or Imprecise Result or Readings	26/04/2021	26/04/2021
							F19	Surgical Intervention	A0908	Incorrect, Inadequate or Imprecise Result or Readings	26/04/2021	26/04/2021
							F1903	Device Explantation	A0908	Incorrect, Inadequate or Imprecise Result or Readings	26/04/2021	26/04/2021
					CT3006	Implantable inorganic material	E2403	No Clinical Signs, Symptoms or Conditions	A0908	Incorrect, Inadequate or Imprecise Result or Readings	26/04/2021	26/04/2021
							F08	Hospitalization or Prolonged Hospitalization	A0908	Incorrect, Inadequate or Imprecise Result or Readings	26/04/2021	26/04/2021
							F19	Surgical Intervention	A0908	Incorrect, Inadequate or Imprecise Result or Readings	26/04/2021	26/04/2021
							F1903	Device Explantation	A0908	Incorrect, Inadequate or Imprecise Result or Readings	26/04/2021	26/04/2021
					CT314	Electrically powered	E2403	No Clinical Signs, Symptoms or Conditions	A0908	Incorrect, Inadequate or Imprecise Result or Readings	26/04/2021	26/04/2021
							F08	Hospitalization or Prolonged Hospitalization	A0908	Incorrect, Inadequate or Imprecise Result or Readings	26/04/2021	26/04/2021
							F19	Surgical Intervention	A0908	Incorrect, Inadequate or Imprecise Result or Readings	26/04/2021	26/04/2021
							F1903	Device Explantation	A0908	Incorrect, Inadequate or Imprecise Result or Readings	26/04/2021	26/04/2021
					CT321	Long-term surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A0908	Incorrect, Inadequate or Imprecise Result or Readings	26/04/2021	26/04/2021

							F08	Hospitalization or Prolonged Hospitalization	A0908	Incorrect, Inadequate or Imprecise Result or Readings	26/04/2021	26/04/2021
							F19	Surgical Intervention	A0908	Incorrect, Inadequate or Imprecise Result or Readings	26/04/2021	26/04/2021
							F1903	Device Explantation	A0908	Incorrect, Inadequate or Imprecise Result or Readings	26/04/2021	26/04/2021
					CT449	Generators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A0908	Incorrect, Inadequate or Imprecise Result or Readings	26/04/2021	26/04/2021
							F08	Hospitalization or Prolonged Hospitalization	A0908	Incorrect, Inadequate or Imprecise Result or Readings	26/04/2021	26/04/2021
							F19	Surgical Intervention	A0908	Incorrect, Inadequate or Imprecise Result or Readings	26/04/2021	26/04/2021
							F1903	Device Explantation	A0908	Incorrect, Inadequate or Imprecise Result or Readings	26/04/2021	26/04/2021
					CT981	Single use	E2403	No Clinical Signs, Symptoms or Conditions	A0908	Incorrect, Inadequate or Imprecise Result or Readings	26/04/2021	26/04/2021
							F08	Hospitalization or Prolonged Hospitalization	A0908	Incorrect, Inadequate or Imprecise Result or Readings	26/04/2021	26/04/2021
							F19	Surgical Intervention	A0908	Incorrect, Inadequate or Imprecise Result or Readings	26/04/2021	26/04/2021
							F1903	Device Explantation	A0908	Incorrect, Inadequate or Imprecise Result or Readings	26/04/2021	26/04/2021
					CT983	Surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A0908	Incorrect, Inadequate or Imprecise Result or Readings	26/04/2021	26/04/2021
							F08	Hospitalization or Prolonged Hospitalization	A0908	Incorrect, Inadequate or Imprecise Result or Readings	26/04/2021	26/04/2021
							F19	Surgical Intervention	A0908	Incorrect, Inadequate or Imprecise Result or Readings	26/04/2021	26/04/2021
							F1903	Device Explantation	A0908	Incorrect, Inadequate or Imprecise Result or Readings	26/04/2021	26/04/2021
10954	2021	All other reportable incidents	47265	Dual-chamber implantable pacemaker, rate-responsive	CT141	Electrodes/Leads and associated devices	E011903	Syncope/Fainting	A070101	Failure to Capture	14/04/2021	14/04/2021
									A0712	Pacing Problem	14/04/2021	14/04/2021
							E2103	Device Overstimulation of Tissue	A070101	Failure to Capture	14/04/2021	14/04/2021
									A0712	Pacing Problem	14/04/2021	14/04/2021
							F08	Hospitalization or Prolonged Hospitalization	A070101	Failure to Capture	14/04/2021	14/04/2021
									A0712	Pacing Problem	14/04/2021	14/04/2021
							F12	Serious Injury/ Illness/ Impairment	A070101	Failure to Capture	14/04/2021	14/04/2021
									A0712	Pacing Problem	14/04/2021	14/04/2021
							F19	Surgical Intervention	A070101	Failure to Capture	14/04/2021	14/04/2021
									A0712	Pacing Problem	14/04/2021	14/04/2021
							F1905	Device Revision or Replacement	A070101	Failure to Capture	14/04/2021	14/04/2021
									A0712	Pacing Problem	14/04/2021	14/04/2021
					CT145	Cardiovascular devices	E011903	Syncope/Fainting	A070101	Failure to Capture	14/04/2021	14/04/2021
									A0712	Pacing Problem	14/04/2021	14/04/2021
							E2103	Device Overstimulation of Tissue	A070101	Failure to Capture	14/04/2021	14/04/2021
									A0712	Pacing Problem	14/04/2021	14/04/2021
							F08	Hospitalization or Prolonged Hospitalization	A070101	Failure to Capture	14/04/2021	14/04/2021
									A0712	Pacing Problem	14/04/2021	14/04/2021
							F12	Serious Injury/ Illness/ Impairment	A070101	Failure to Capture	14/04/2021	14/04/2021
									A0712	Pacing Problem	14/04/2021	14/04/2021
							F19	Surgical Intervention	A070101	Failure to Capture	14/04/2021	14/04/2021
									A0712	Pacing Problem	14/04/2021	14/04/2021
							F1905	Device Revision or Replacement	A070101	Failure to Capture	14/04/2021	14/04/2021
									A0712	Pacing Problem	14/04/2021	14/04/2021
					CT1684	Cardiac pacemakers and associated devices	E011903	Syncope/Fainting	A070101	Failure to Capture	14/04/2021	14/04/2021
									A0712	Pacing Problem	14/04/2021	14/04/2021
							E2103	Device Overstimulation of Tissue	A070101	Failure to Capture	14/04/2021	14/04/2021
									A0712	Pacing Problem	14/04/2021	14/04/2021
							F08	Hospitalization or Prolonged Hospitalization	A070101	Failure to Capture	14/04/2021	14/04/2021
									A0712	Pacing Problem	14/04/2021	14/04/2021
							F12	Serious Injury/ Illness/ Impairment	A070101	Failure to Capture	14/04/2021	14/04/2021
									A0712	Pacing Problem	14/04/2021	14/04/2021
							F19	Surgical Intervention	A070101	Failure to Capture	14/04/2021	14/04/2021
									A0712	Pacing Problem	14/04/2021	14/04/2021
							F1905	Device Revision or Replacement	A070101	Failure to Capture	14/04/2021	14/04/2021
									A0712	Pacing Problem	14/04/2021	14/04/2021
					CT1685	Cardiac pacemakers	E011903	Syncope/Fainting	A070101	Failure to Capture	14/04/2021	14/04/2021
									A0712	Pacing Problem	14/04/2021	14/04/2021
							E2103	Device Overstimulation of Tissue	A070101	Failure to Capture	14/04/2021	14/04/2021
									A0712	Pacing Problem	14/04/2021	14/04/2021
							F08	Hospitalization or Prolonged Hospitalization	A070101	Failure to Capture	14/04/2021	14/04/2021
									A0712	Pacing Problem	14/04/2021	14/04/2021
							F12	Serious Injury/ Illness/ Impairment	A070101	Failure to Capture	14/04/2021	14/04/2021
									A0712	Pacing Problem	14/04/2021	14/04/2021
							F19	Surgical Intervention	A070101	Failure to Capture	14/04/2021	14/04/2021
									A0712	Pacing Problem	14/04/2021	14/04/2021
							F1905	Device Revision or Replacement	A070101	Failure to Capture	14/04/2021	14/04/2021
									A0712	Pacing Problem	14/04/2021	14/04/2021
					CT233	Surgical	E011903	Syncope/Fainting	A070101	Failure to Capture	14/04/2021	14/04/2021
									A0712	Pacing Problem	14/04/2021	14/04/2021
							E2103	Device Overstimulation of Tissue	A070101	Failure to Capture	14/04/2021	14/04/2021
									A0712	Pacing Problem	14/04/2021	14/04/2021
							F08	Hospitalization or Prolonged Hospitalization	A070101	Failure to Capture	14/04/2021	14/04/2021
									A0712	Pacing Problem	14/04/2021	14/04/2021
							F12	Serious Injury/ Illness/ Impairment	A070101	Failure to Capture	14/04/2021	14/04/2021
									A0712	Pacing Problem	14/04/2021	14/04/2021
							F19	Surgical Intervention	A070101	Failure to Capture	14/04/2021	14/04/2021
									A0712	Pacing Problem	14/04/2021	14/04/2021
							F1905	Device Revision or Replacement	A070101	Failure to Capture	14/04/2021	14/04/2021
									A0712	Pacing Problem	14/04/2021	14/04/2021
					CT2406	Implantable	E011903	Syncope/Fainting	A070101	Failure to Capture	14/04/2021	14/04/2021

									A0712	Pacing Problem	14/04/2021	14/04/2021
							E2103	Device Overstimulation of Tissue	A070101	Failure to Capture	14/04/2021	14/04/2021
									A0712	Pacing Problem	14/04/2021	14/04/2021
							F08	Hospitalization or Prolonged Hospitalization	A070101	Failure to Capture	14/04/2021	14/04/2021
									A0712	Pacing Problem	14/04/2021	14/04/2021
							F12	Serious Injury/ Illness/ Impairment	A070101	Failure to Capture	14/04/2021	14/04/2021
									A0712	Pacing Problem	14/04/2021	14/04/2021
							F19	Surgical Intervention	A070101	Failure to Capture	14/04/2021	14/04/2021
									A0712	Pacing Problem	14/04/2021	14/04/2021
							F1905	Device Revision or Replacement	A070101	Failure to Capture	14/04/2021	14/04/2021
									A0712	Pacing Problem	14/04/2021	14/04/2021
					CT2407	Active implantable	E011903	Syncope/Fainting	A070101	Failure to Capture	14/04/2021	14/04/2021
									A0712	Pacing Problem	14/04/2021	14/04/2021
							E2103	Device Overstimulation of Tissue	A070101	Failure to Capture	14/04/2021	14/04/2021
									A0712	Pacing Problem	14/04/2021	14/04/2021
							F08	Hospitalization or Prolonged Hospitalization	A070101	Failure to Capture	14/04/2021	14/04/2021
									A0712	Pacing Problem	14/04/2021	14/04/2021
							F12	Serious Injury/ Illness/ Impairment	A070101	Failure to Capture	14/04/2021	14/04/2021
									A0712	Pacing Problem	14/04/2021	14/04/2021
							F19	Surgical Intervention	A070101	Failure to Capture	14/04/2021	14/04/2021
									A0712	Pacing Problem	14/04/2021	14/04/2021
							F1905	Device Revision or Replacement	A070101	Failure to Capture	14/04/2021	14/04/2021
									A0712	Pacing Problem	14/04/2021	14/04/2021
					CT257	Implantable cardiac pacemakers	E011903	Syncope/Fainting	A070101	Failure to Capture	14/04/2021	14/04/2021
									A0712	Pacing Problem	14/04/2021	14/04/2021
							E2103	Device Overstimulation of Tissue	A070101	Failure to Capture	14/04/2021	14/04/2021
									A0712	Pacing Problem	14/04/2021	14/04/2021
							F08	Hospitalization or Prolonged Hospitalization	A070101	Failure to Capture	14/04/2021	14/04/2021
									A0712	Pacing Problem	14/04/2021	14/04/2021
							F12	Serious Injury/ Illness/ Impairment	A070101	Failure to Capture	14/04/2021	14/04/2021
									A0712	Pacing Problem	14/04/2021	14/04/2021
							F19	Surgical Intervention	A070101	Failure to Capture	14/04/2021	14/04/2021
									A0712	Pacing Problem	14/04/2021	14/04/2021
							F1905	Device Revision or Replacement	A070101	Failure to Capture	14/04/2021	14/04/2021
									A0712	Pacing Problem	14/04/2021	14/04/2021
					CT2727	Active implant devices/ancillaries	E011903	Syncope/Fainting	A070101	Failure to Capture	14/04/2021	14/04/2021
									A0712	Pacing Problem	14/04/2021	14/04/2021
							E2103	Device Overstimulation of Tissue	A070101	Failure to Capture	14/04/2021	14/04/2021
									A0712	Pacing Problem	14/04/2021	14/04/2021
							F08	Hospitalization or Prolonged Hospitalization	A070101	Failure to Capture	14/04/2021	14/04/2021
									A0712	Pacing Problem	14/04/2021	14/04/2021
							F12	Serious Injury/ Illness/ Impairment	A070101	Failure to Capture	14/04/2021	14/04/2021
									A0712	Pacing Problem	14/04/2021	14/04/2021
							F19	Surgical Intervention	A070101	Failure to Capture	14/04/2021	14/04/2021
									A0712	Pacing Problem	14/04/2021	14/04/2021
							F1905	Device Revision or Replacement	A070101	Failure to Capture	14/04/2021	14/04/2021
									A0712	Pacing Problem	14/04/2021	14/04/2021
					CT2750	Implantable cardiac electrical stimulators and associated devices	E011903	Syncope/Fainting	A070101	Failure to Capture	14/04/2021	14/04/2021
									A0712	Pacing Problem	14/04/2021	14/04/2021
							E2103	Device Overstimulation of Tissue	A070101	Failure to Capture	14/04/2021	14/04/2021
									A0712	Pacing Problem	14/04/2021	14/04/2021
							F08	Hospitalization or Prolonged Hospitalization	A070101	Failure to Capture	14/04/2021	14/04/2021
									A0712	Pacing Problem	14/04/2021	14/04/2021
							F12	Serious Injury/ Illness/ Impairment	A070101	Failure to Capture	14/04/2021	14/04/2021
									A0712	Pacing Problem	14/04/2021	14/04/2021
							F19	Surgical Intervention	A070101	Failure to Capture	14/04/2021	14/04/2021
									A0712	Pacing Problem	14/04/2021	14/04/2021
							F1905	Device Revision or Replacement	A070101	Failure to Capture	14/04/2021	14/04/2021
									A0712	Pacing Problem	14/04/2021	14/04/2021
					CT2866	Implantable cardiac electrical stimulators	E011903	Syncope/Fainting	A070101	Failure to Capture	14/04/2021	14/04/2021
									A0712	Pacing Problem	14/04/2021	14/04/2021
							E2103	Device Overstimulation of Tissue	A070101	Failure to Capture	14/04/2021	14/04/2021
									A0712	Pacing Problem	14/04/2021	14/04/2021
							F08	Hospitalization or Prolonged Hospitalization	A070101	Failure to Capture	14/04/2021	14/04/2021
									A0712	Pacing Problem	14/04/2021	14/04/2021
							F12	Serious Injury/ Illness/ Impairment	A070101	Failure to Capture	14/04/2021	14/04/2021
									A0712	Pacing Problem	14/04/2021	14/04/2021
							F19	Surgical Intervention	A070101	Failure to Capture	14/04/2021	14/04/2021
									A0712	Pacing Problem	14/04/2021	14/04/2021
							F1905	Device Revision or Replacement	A070101	Failure to Capture	14/04/2021	14/04/2021
									A0712	Pacing Problem	14/04/2021	14/04/2021
					CT2985	Active	E011903	Syncope/Fainting	A070101	Failure to Capture	14/04/2021	14/04/2021
									A0712	Pacing Problem	14/04/2021	14/04/2021
							E2103	Device Overstimulation of Tissue	A070101	Failure to Capture	14/04/2021	14/04/2021
									A0712	Pacing Problem	14/04/2021	14/04/2021
							F08	Hospitalization or Prolonged Hospitalization	A070101	Failure to Capture	14/04/2021	14/04/2021
									A0712	Pacing Problem	14/04/2021	14/04/2021
							F12	Serious Injury/ Illness/ Impairment	A070101	Failure to Capture	14/04/2021	14/04/2021
									A0712	Pacing Problem	14/04/2021	14/04/2021
							F19	Surgical Intervention	A070101	Failure to Capture	14/04/2021	14/04/2021
									A0712	Pacing Problem	14/04/2021	14/04/2021
							F1905	Device Revision or Replacement	A070101	Failure to Capture	14/04/2021	14/04/2021
									A0712	Pacing Problem	14/04/2021	14/04/2021
					CT3001	Implantable metal	E011903	Syncope/Fainting	A070101	Failure to Capture	14/04/2021	14/04/2021
									A0712	Pacing Problem	14/04/2021	14/04/2021
							E2103	Device Overstimulation of Tissue	A070101	Failure to Capture	14/04/2021	14/04/2021
									A0712	Pacing Problem	14/04/2021	14/04/2021
							F08	Hospitalization or Prolonged Hospitalization	A070101	Failure to Capture	14/04/2021	14/04/2021

									A0712	Pacing Problem	14/04/2021	14/04/2021
							F12	Serious Injury/ Illness/ Impairment	A070101	Failure to Capture	14/04/2021	14/04/2021
									A0712	Pacing Problem	14/04/2021	14/04/2021
							F19	Surgical Intervention	A070101	Failure to Capture	14/04/2021	14/04/2021
									A0712	Pacing Problem	14/04/2021	14/04/2021
							F1905	Device Revision or Replacement	A070101	Failure to Capture	14/04/2021	14/04/2021
									A0712	Pacing Problem	14/04/2021	14/04/2021
					CT3002	Implantable synthetic polymer	E011903	Syncope/Fainting	A070101	Failure to Capture	14/04/2021	14/04/2021
									A0712	Pacing Problem	14/04/2021	14/04/2021
							E2103	Device Overstimulation of Tissue	A070101	Failure to Capture	14/04/2021	14/04/2021
									A0712	Pacing Problem	14/04/2021	14/04/2021
							F08	Hospitalization or Prolonged Hospitalization	A070101	Failure to Capture	14/04/2021	14/04/2021
									A0712	Pacing Problem	14/04/2021	14/04/2021
							F12	Serious Injury/ Illness/ Impairment	A070101	Failure to Capture	14/04/2021	14/04/2021
									A0712	Pacing Problem	14/04/2021	14/04/2021
							F19	Surgical Intervention	A070101	Failure to Capture	14/04/2021	14/04/2021
									A0712	Pacing Problem	14/04/2021	14/04/2021
							F1905	Device Revision or Replacement	A070101	Failure to Capture	14/04/2021	14/04/2021
									A0712	Pacing Problem	14/04/2021	14/04/2021
					CT3006	Implantable inorganic material	E011903	Syncope/Fainting	A070101	Failure to Capture	14/04/2021	14/04/2021
									A0712	Pacing Problem	14/04/2021	14/04/2021
							E2103	Device Overstimulation of Tissue	A070101	Failure to Capture	14/04/2021	14/04/2021
									A0712	Pacing Problem	14/04/2021	14/04/2021
							F08	Hospitalization or Prolonged Hospitalization	A070101	Failure to Capture	14/04/2021	14/04/2021
									A0712	Pacing Problem	14/04/2021	14/04/2021
							F12	Serious Injury/ Illness/ Impairment	A070101	Failure to Capture	14/04/2021	14/04/2021
									A0712	Pacing Problem	14/04/2021	14/04/2021
							F19	Surgical Intervention	A070101	Failure to Capture	14/04/2021	14/04/2021
									A0712	Pacing Problem	14/04/2021	14/04/2021
							F1905	Device Revision or Replacement	A070101	Failure to Capture	14/04/2021	14/04/2021
									A0712	Pacing Problem	14/04/2021	14/04/2021
					CT314	Electrically powered	E011903	Syncope/Fainting	A070101	Failure to Capture	14/04/2021	14/04/2021
									A0712	Pacing Problem	14/04/2021	14/04/2021
							E2103	Device Overstimulation of Tissue	A070101	Failure to Capture	14/04/2021	14/04/2021
									A0712	Pacing Problem	14/04/2021	14/04/2021
							F08	Hospitalization or Prolonged Hospitalization	A070101	Failure to Capture	14/04/2021	14/04/2021
									A0712	Pacing Problem	14/04/2021	14/04/2021
							F12	Serious Injury/ Illness/ Impairment	A070101	Failure to Capture	14/04/2021	14/04/2021
									A0712	Pacing Problem	14/04/2021	14/04/2021
							F19	Surgical Intervention	A070101	Failure to Capture	14/04/2021	14/04/2021
									A0712	Pacing Problem	14/04/2021	14/04/2021
							F1905	Device Revision or Replacement	A070101	Failure to Capture	14/04/2021	14/04/2021
									A0712	Pacing Problem	14/04/2021	14/04/2021
					CT321	Long-term surgical invasive	E011903	Syncope/Fainting	A070101	Failure to Capture	14/04/2021	14/04/2021
									A0712	Pacing Problem	14/04/2021	14/04/2021
							E2103	Device Overstimulation of Tissue	A070101	Failure to Capture	14/04/2021	14/04/2021
									A0712	Pacing Problem	14/04/2021	14/04/2021
							F08	Hospitalization or Prolonged Hospitalization	A070101	Failure to Capture	14/04/2021	14/04/2021
									A0712	Pacing Problem	14/04/2021	14/04/2021
							F12	Serious Injury/ Illness/ Impairment	A070101	Failure to Capture	14/04/2021	14/04/2021
									A0712	Pacing Problem	14/04/2021	14/04/2021
							F19	Surgical Intervention	A070101	Failure to Capture	14/04/2021	14/04/2021
									A0712	Pacing Problem	14/04/2021	14/04/2021
							F1905	Device Revision or Replacement	A070101	Failure to Capture	14/04/2021	14/04/2021
									A0712	Pacing Problem	14/04/2021	14/04/2021
					CT449	Generators and associated devices	E011903	Syncope/Fainting	A070101	Failure to Capture	14/04/2021	14/04/2021
									A0712	Pacing Problem	14/04/2021	14/04/2021
							E2103	Device Overstimulation of Tissue	A070101	Failure to Capture	14/04/2021	14/04/2021
									A0712	Pacing Problem	14/04/2021	14/04/2021
							F08	Hospitalization or Prolonged Hospitalization	A070101	Failure to Capture	14/04/2021	14/04/2021
									A0712	Pacing Problem	14/04/2021	14/04/2021
							F12	Serious Injury/ Illness/ Impairment	A070101	Failure to Capture	14/04/2021	14/04/2021
									A0712	Pacing Problem	14/04/2021	14/04/2021
							F19	Surgical Intervention	A070101	Failure to Capture	14/04/2021	14/04/2021
									A0712	Pacing Problem	14/04/2021	14/04/2021
							F1905	Device Revision or Replacement	A070101	Failure to Capture	14/04/2021	14/04/2021
									A0712	Pacing Problem	14/04/2021	14/04/2021
					CT981	Single use	E011903	Syncope/Fainting	A070101	Failure to Capture	14/04/2021	14/04/2021
									A0712	Pacing Problem	14/04/2021	14/04/2021
							E2103	Device Overstimulation of Tissue	A070101	Failure to Capture	14/04/2021	14/04/2021
									A0712	Pacing Problem	14/04/2021	14/04/2021
							F08	Hospitalization or Prolonged Hospitalization	A070101	Failure to Capture	14/04/2021	14/04/2021
									A0712	Pacing Problem	14/04/2021	14/04/2021
							F12	Serious Injury/ Illness/ Impairment	A070101	Failure to Capture	14/04/2021	14/04/2021
									A0712	Pacing Problem	14/04/2021	14/04/2021
							F19	Surgical Intervention	A070101	Failure to Capture	14/04/2021	14/04/2021
									A0712	Pacing Problem	14/04/2021	14/04/2021
							F1905	Device Revision or Replacement	A070101	Failure to Capture	14/04/2021	14/04/2021
									A0712	Pacing Problem	14/04/2021	14/04/2021
					CT983	Surgical invasive	E011903	Syncope/Fainting	A070101	Failure to Capture	14/04/2021	14/04/2021
									A0712	Pacing Problem	14/04/2021	14/04/2021
							E2103	Device Overstimulation of Tissue	A070101	Failure to Capture	14/04/2021	14/04/2021
									A0712	Pacing Problem	14/04/2021	14/04/2021
							F08	Hospitalization or Prolonged Hospitalization	A070101	Failure to Capture	14/04/2021	14/04/2021
									A0712	Pacing Problem	14/04/2021	14/04/2021
							F12	Serious Injury/ Illness/ Impairment	A070101	Failure to Capture	14/04/2021	14/04/2021
									A0712	Pacing Problem	14/04/2021	14/04/2021
							F19	Surgical Intervention	A070101	Failure to Capture	14/04/2021	14/04/2021

									A0712	Pacing Problem	14/04/2021	14/04/2021
							F1905	Device Revision or Replacement	A070101	Failure to Capture	14/04/2021	14/04/2021
									A0712	Pacing Problem	14/04/2021	14/04/2021
10955	2021	All other reportable incidents	47265	Dual-chamber implantable pacemaker, rate-responsive	CT141	Electrodes/Leads and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	23/05/2021	23/05/2021
									A1304	Telemetry Discrepancy	23/05/2021	23/05/2021
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	23/05/2021	23/05/2021
									A1304	Telemetry Discrepancy	23/05/2021	23/05/2021
							F19	Surgical Intervention	A070504	Premature Discharge of Battery	23/05/2021	23/05/2021
									A1304	Telemetry Discrepancy	23/05/2021	23/05/2021
					CT145	Cardiovascular devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	23/05/2021	23/05/2021
									A1304	Telemetry Discrepancy	23/05/2021	23/05/2021
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	23/05/2021	23/05/2021
									A1304	Telemetry Discrepancy	23/05/2021	23/05/2021
							F19	Surgical Intervention	A070504	Premature Discharge of Battery	23/05/2021	23/05/2021
									A1304	Telemetry Discrepancy	23/05/2021	23/05/2021
					CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	23/05/2021	23/05/2021
									A1304	Telemetry Discrepancy	23/05/2021	23/05/2021
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	23/05/2021	23/05/2021
									A1304	Telemetry Discrepancy	23/05/2021	23/05/2021
							F19	Surgical Intervention	A070504	Premature Discharge of Battery	23/05/2021	23/05/2021
									A1304	Telemetry Discrepancy	23/05/2021	23/05/2021
					CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	23/05/2021	23/05/2021
									A1304	Telemetry Discrepancy	23/05/2021	23/05/2021
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	23/05/2021	23/05/2021
									A1304	Telemetry Discrepancy	23/05/2021	23/05/2021
							F19	Surgical Intervention	A070504	Premature Discharge of Battery	23/05/2021	23/05/2021
									A1304	Telemetry Discrepancy	23/05/2021	23/05/2021
					CT233	Surgical	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	23/05/2021	23/05/2021
									A1304	Telemetry Discrepancy	23/05/2021	23/05/2021
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	23/05/2021	23/05/2021
									A1304	Telemetry Discrepancy	23/05/2021	23/05/2021
							F19	Surgical Intervention	A070504	Premature Discharge of Battery	23/05/2021	23/05/2021
									A1304	Telemetry Discrepancy	23/05/2021	23/05/2021
					CT2406	Implantable	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	23/05/2021	23/05/2021
									A1304	Telemetry Discrepancy	23/05/2021	23/05/2021
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	23/05/2021	23/05/2021
									A1304	Telemetry Discrepancy	23/05/2021	23/05/2021
							F19	Surgical Intervention	A070504	Premature Discharge of Battery	23/05/2021	23/05/2021
									A1304	Telemetry Discrepancy	23/05/2021	23/05/2021
					CT2407	Active implantable	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	23/05/2021	23/05/2021
									A1304	Telemetry Discrepancy	23/05/2021	23/05/2021
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	23/05/2021	23/05/2021
									A1304	Telemetry Discrepancy	23/05/2021	23/05/2021
							F19	Surgical Intervention	A070504	Premature Discharge of Battery	23/05/2021	23/05/2021
									A1304	Telemetry Discrepancy	23/05/2021	23/05/2021
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	23/05/2021	23/05/2021
									A1304	Telemetry Discrepancy	23/05/2021	23/05/2021
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	23/05/2021	23/05/2021
									A1304	Telemetry Discrepancy	23/05/2021	23/05/2021
							F19	Surgical Intervention	A070504	Premature Discharge of Battery	23/05/2021	23/05/2021
									A1304	Telemetry Discrepancy	23/05/2021	23/05/2021
					CT2727	Active implant devices/ancillaries	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	23/05/2021	23/05/2021
									A1304	Telemetry Discrepancy	23/05/2021	23/05/2021
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	23/05/2021	23/05/2021
									A1304	Telemetry Discrepancy	23/05/2021	23/05/2021
							F19	Surgical Intervention	A070504	Premature Discharge of Battery	23/05/2021	23/05/2021
									A1304	Telemetry Discrepancy	23/05/2021	23/05/2021
					CT2750	Implantable cardiac electrical stimulators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	23/05/2021	23/05/2021
									A1304	Telemetry Discrepancy	23/05/2021	23/05/2021
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	23/05/2021	23/05/2021
									A1304	Telemetry Discrepancy	23/05/2021	23/05/2021
							F19	Surgical Intervention	A070504	Premature Discharge of Battery	23/05/2021	23/05/2021
									A1304	Telemetry Discrepancy	23/05/2021	23/05/2021
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	23/05/2021	23/05/2021
									A1304	Telemetry Discrepancy	23/05/2021	23/05/2021
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	23/05/2021	23/05/2021
									A1304	Telemetry Discrepancy	23/05/2021	23/05/2021
							F19	Surgical Intervention	A070504	Premature Discharge of Battery	23/05/2021	23/05/2021
									A1304	Telemetry Discrepancy	23/05/2021	23/05/2021
					CT2985	Active	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	23/05/2021	23/05/2021
									A1304	Telemetry Discrepancy	23/05/2021	23/05/2021
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	23/05/2021	23/05/2021
									A1304	Telemetry Discrepancy	23/05/2021	23/05/2021
							F19	Surgical Intervention	A070504	Premature Discharge of Battery	23/05/2021	23/05/2021
									A1304	Telemetry Discrepancy	23/05/2021	23/05/2021
					CT3001	Implantable metal	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	23/05/2021	23/05/2021
									A1304	Telemetry Discrepancy	23/05/2021	23/05/2021
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	23/05/2021	23/05/2021
									A1304	Telemetry Discrepancy	23/05/2021	23/05/2021
							F19	Surgical Intervention	A070504	Premature Discharge of Battery	23/05/2021	23/05/2021
									A1304	Telemetry Discrepancy	23/05/2021	23/05/2021
					CT3002	Implantable synthetic polymer	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	23/05/2021	23/05/2021
									A1304	Telemetry Discrepancy	23/05/2021	23/05/2021
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	23/05/2021	23/05/2021
									A1304	Telemetry Discrepancy	23/05/2021	23/05/2021
							F19	Surgical Intervention	A070504	Premature Discharge of Battery	23/05/2021	23/05/2021
									A1304	Telemetry Discrepancy	23/05/2021	23/05/2021
					CT3006	Implantable inorganic material	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	23/05/2021	23/05/2021

								A1304	Telemetry Discrepancy	23/05/2021	23/05/2021	
						F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	23/05/2021	23/05/2021	
								A1304	Telemetry Discrepancy	23/05/2021	23/05/2021	
						F19	Surgical Intervention	A070504	Premature Discharge of Battery	23/05/2021	23/05/2021	
								A1304	Telemetry Discrepancy	23/05/2021	23/05/2021	
				CT314	Electrically powered	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	23/05/2021	23/05/2021	
								A1304	Telemetry Discrepancy	23/05/2021	23/05/2021	
						F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	23/05/2021	23/05/2021	
								A1304	Telemetry Discrepancy	23/05/2021	23/05/2021	
						F19	Surgical Intervention	A070504	Premature Discharge of Battery	23/05/2021	23/05/2021	
								A1304	Telemetry Discrepancy	23/05/2021	23/05/2021	
				CT321	Long-term surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	23/05/2021	23/05/2021	
								A1304	Telemetry Discrepancy	23/05/2021	23/05/2021	
						F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	23/05/2021	23/05/2021	
								A1304	Telemetry Discrepancy	23/05/2021	23/05/2021	
						F19	Surgical Intervention	A070504	Premature Discharge of Battery	23/05/2021	23/05/2021	
								A1304	Telemetry Discrepancy	23/05/2021	23/05/2021	
				CT449	Generators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	23/05/2021	23/05/2021	
								A1304	Telemetry Discrepancy	23/05/2021	23/05/2021	
						F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	23/05/2021	23/05/2021	
								A1304	Telemetry Discrepancy	23/05/2021	23/05/2021	
						F19	Surgical Intervention	A070504	Premature Discharge of Battery	23/05/2021	23/05/2021	
								A1304	Telemetry Discrepancy	23/05/2021	23/05/2021	
				CT981	Single use	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	23/05/2021	23/05/2021	
								A1304	Telemetry Discrepancy	23/05/2021	23/05/2021	
						F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	23/05/2021	23/05/2021	
								A1304	Telemetry Discrepancy	23/05/2021	23/05/2021	
						F19	Surgical Intervention	A070504	Premature Discharge of Battery	23/05/2021	23/05/2021	
								A1304	Telemetry Discrepancy	23/05/2021	23/05/2021	
				CT983	Surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	23/05/2021	23/05/2021	
								A1304	Telemetry Discrepancy	23/05/2021	23/05/2021	
						F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	23/05/2021	23/05/2021	
								A1304	Telemetry Discrepancy	23/05/2021	23/05/2021	
						F19	Surgical Intervention	A070504	Premature Discharge of Battery	23/05/2021	23/05/2021	
								A1304	Telemetry Discrepancy	23/05/2021	23/05/2021	
10956	2021	All other reportable incidents	63188	Left-ventricular wireless pacing system sensor/transmitter	CT141	Electrodes/Leads and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070801	Complete Loss of Power		
					CT145	Cardiovascular devices	E2403	No Clinical Signs, Symptoms or Conditions	A070801	Complete Loss of Power		
					CT1577	Wireless monitoring system transmitters	E2403	No Clinical Signs, Symptoms or Conditions	A070801	Complete Loss of Power		
					CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070801	Complete Loss of Power		
					CT2317	Transmitters/Receivers/Transceivers	E2403	No Clinical Signs, Symptoms or Conditions	A070801	Complete Loss of Power		
					CT233	Surgical	E2403	No Clinical Signs, Symptoms or Conditions	A070801	Complete Loss of Power		
					CT2406	Implantable	E2403	No Clinical Signs, Symptoms or Conditions	A070801	Complete Loss of Power		
					CT2407	Active implantable	E2403	No Clinical Signs, Symptoms or Conditions	A070801	Complete Loss of Power		
					CT255	Transducers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070801	Complete Loss of Power		
					CT2726	Physiological monitoring/measuring devices	E2403	No Clinical Signs, Symptoms or Conditions	A070801	Complete Loss of Power		
					CT2727	Active implant devices/ancillaries	E2403	No Clinical Signs, Symptoms or Conditions	A070801	Complete Loss of Power		
					CT2750	Implantable cardiac electrical stimulators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070801	Complete Loss of Power		
					CT2885	Physiological monitoring/measuring system transmitters/receivers/transceivers	E2403	No Clinical Signs, Symptoms or Conditions	A070801	Complete Loss of Power		
					CT2985	Active	E2403	No Clinical Signs, Symptoms or Conditions	A070801	Complete Loss of Power		
					CT3001	Implantable metal	E2403	No Clinical Signs, Symptoms or Conditions	A070801	Complete Loss of Power		
					CT3002	Implantable synthetic polymer	E2403	No Clinical Signs, Symptoms or Conditions	A070801	Complete Loss of Power		
					CT3006	Implantable inorganic material	E2403	No Clinical Signs, Symptoms or Conditions	A070801	Complete Loss of Power		
					CT314	Electrically powered	E2403	No Clinical Signs, Symptoms or Conditions	A070801	Complete Loss of Power		
					CT321	Long-term surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A070801	Complete Loss of Power		
					CT384	Sensors and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070801	Complete Loss of Power		
					CT449	Generators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070801	Complete Loss of Power		
					CT981	Single use	E2403	No Clinical Signs, Symptoms or Conditions	A070801	Complete Loss of Power		
					CT983	Surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A070801	Complete Loss of Power		
					CT994	Sensors	E2403	No Clinical Signs, Symptoms or Conditions	A070801	Complete Loss of Power		
10957	2021	All other reportable incidents	63188	Left-ventricular wireless pacing system sensor/transmitter	CT141	Electrodes/Leads and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070801	Complete Loss of Power		
					CT145	Cardiovascular devices	E2403	No Clinical Signs, Symptoms or Conditions	A070801	Complete Loss of Power		
					CT1577	Wireless monitoring system transmitters	E2403	No Clinical Signs, Symptoms or Conditions	A070801	Complete Loss of Power		
					CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070801	Complete Loss of Power		
					CT2317	Transmitters/Receivers/Transceivers	E2403	No Clinical Signs, Symptoms or Conditions	A070801	Complete Loss of Power		
					CT233	Surgical	E2403	No Clinical Signs, Symptoms or Conditions	A070801	Complete Loss of Power		
					CT2406	Implantable	E2403	No Clinical Signs, Symptoms or Conditions	A070801	Complete Loss of Power		
					CT2407	Active implantable	E2403	No Clinical Signs, Symptoms or Conditions	A070801	Complete Loss of Power		
					CT255	Transducers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070801	Complete Loss of Power		
					CT2726	Physiological monitoring/measuring devices	E2403	No Clinical Signs, Symptoms or Conditions	A070801	Complete Loss of Power		
					CT2727	Active implant devices/ancillaries	E2403	No Clinical Signs, Symptoms or Conditions	A070801	Complete Loss of Power		
					CT2750	Implantable cardiac electrical stimulators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070801	Complete Loss of Power		
					CT2885	Physiological monitoring/measuring system transmitters/receivers/transceivers	E2403	No Clinical Signs, Symptoms or Conditions	A070801	Complete Loss of Power		
					CT2985	Active	E2403	No Clinical Signs, Symptoms or Conditions	A070801	Complete Loss of Power		
					CT3001	Implantable metal	E2403	No Clinical Signs, Symptoms or Conditions	A070801	Complete Loss of Power		
					CT3002	Implantable synthetic polymer	E2403	No Clinical Signs, Symptoms or Conditions	A070801	Complete Loss of Power		
					CT3006	Implantable inorganic material	E2403	No Clinical Signs, Symptoms or Conditions	A070801	Complete Loss of Power		
					CT314	Electrically powered	E2403	No Clinical Signs, Symptoms or Conditions	A070801	Complete Loss of Power		
					CT321	Long-term surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A070801	Complete Loss of Power		
					CT384	Sensors and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070801	Complete Loss of Power		
					CT449	Generators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070801	Complete Loss of Power		
					CT981	Single use	E2403	No Clinical Signs, Symptoms or Conditions	A070801	Complete Loss of Power		
					CT983	Surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A070801	Complete Loss of Power		
					CT994	Sensors	E2403	No Clinical Signs, Symptoms or Conditions	A070801	Complete Loss of Power		
10958	2021	All other reportable incidents	37265	Dual-chamber implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		

10959	2021	All other reportable incidents	35852	Single-chamber implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A0709	Device Sensing Problem		
									A070909	Oversensing		
							F15	Recognized Device or Procedural Complication	A0709	Device Sensing Problem		
									A070909	Oversensing		
							F26	No Health Consequences or Impact	A0709	Device Sensing Problem		
									A070909	Oversensing		
10960	2021	All other reportable incidents	37265	Dual-chamber implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A070909	Oversensing		
									A0712	Pacing Problem		
							F26	No Health Consequences or Impact	A070909	Oversensing		
									A0712	Pacing Problem		
10961	2021	All other reportable incidents	17882	Physiologic-monitoring defibrillation system	CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A0719	Unexpected Shutdown		
							F26	No Health Consequences or Impact	A0719	Unexpected Shutdown		
10962	2021	All other reportable incidents	35852	Single-chamber implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A25	No Apparent Adverse Event	16/03/2021	16/03/2021
							F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A25	No Apparent Adverse Event	16/03/2021	16/03/2021
							F1905	Device Revision or Replacement	A25	No Apparent Adverse Event	16/03/2021	16/03/2021
10963	2021	All other reportable incidents	35852	Single-chamber implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E2401	Insufficient Information	A16	Protective Measures Problem	24/03/2021	24/03/2021
							F1905	Device Revision or Replacement	A16	Protective Measures Problem	24/03/2021	24/03/2021
10964	2021	All other reportable incidents	47270	Cardiac resynchronization therapy implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A070909	Oversensing		
									A071302	Inappropriate/Inadequate Shock/Stimulation		
							F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A070909	Oversensing		
									A071302	Inappropriate/Inadequate Shock/Stimulation		
							F26	No Health Consequences or Impact	A070909	Oversensing		
									A071302	Inappropriate/Inadequate Shock/Stimulation		
10965	2021	All other reportable incidents	37265	Dual-chamber implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A070909	Oversensing		
									A0712	Pacing Problem		
							F26	No Health Consequences or Impact	A070909	Oversensing		
									A0712	Pacing Problem		
10966	2021	All other reportable incidents	47265	Dual-chamber implantable pacemaker, rate-responsive	CT141	Electrodes/Leads and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT145	Cardiovascular devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT233	Surgical	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT2406	Implantable	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT2407	Active implantable	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT2727	Active implant devices/ancillaries	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT2750	Implantable cardiac electrical stimulators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT2985	Active	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT3001	Implantable metal	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT3002	Implantable synthetic polymer	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT3006	Implantable inorganic material	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT314	Electrically powered	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT321	Long-term surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT449	Generators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT981	Single use	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT983	Surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
10967	2021	All other reportable incidents	35852	Single-chamber implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	28/04/2021	28/04/2021
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	28/04/2021	28/04/2021
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	28/04/2021	28/04/2021
10968	2021	All other reportable incidents	17882	Physiologic-monitoring defibrillation system	CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A0719	Unexpected Shutdown		
							F26	No Health Consequences or Impact	A0719	Unexpected Shutdown		
10969	2021	All other reportable incidents	47263	Cardiac resynchronization therapy implantable pacemaker	CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	05/05/2021	05/05/2021
							F1901	Additional Surgery	A070504	Premature Discharge of Battery	05/05/2021	05/05/2021
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	05/05/2021	05/05/2021
					CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	05/05/2021	05/05/2021
							F1901	Additional Surgery	A070504	Premature Discharge of Battery	05/05/2021	05/05/2021
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	05/05/2021	05/05/2021
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	05/05/2021	05/05/2021
							F1901	Additional Surgery	A070504	Premature Discharge of Battery	05/05/2021	05/05/2021
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	05/05/2021	05/05/2021
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	05/05/2021	05/05/2021
							F1901	Additional Surgery	A070504	Premature Discharge of Battery	05/05/2021	05/05/2021
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	05/05/2021	05/05/2021
10970	2021	All other reportable incidents	36241	Pacemaker/defibrillator lead	CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A072201	High Impedance		
							F26	No Health Consequences or Impact	A072201	High Impedance		
10971	2021	All other reportable incidents	47265	Dual-chamber implantable pacemaker, rate-responsive	CT141	Electrodes/Leads and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem	08/06/2021	08/06/2021
									A070504	Premature Discharge of Battery	08/06/2021	08/06/2021

				F08	Hospitalization or Prolonged Hospitalization	A0705	Battery Problem	08/06/2021	08/06/2021
						A070504	Premature Discharge of Battery	08/06/2021	08/06/2021
				F1903	Device Explantation	A0705	Battery Problem	08/06/2021	08/06/2021
						A070504	Premature Discharge of Battery	08/06/2021	08/06/2021
				CT145	Cardiovascular devices	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem	08/06/2021	08/06/2021
								A070504	Premature Discharge of Battery	08/06/2021	08/06/2021
						F08	Hospitalization or Prolonged Hospitalization	A0705	Battery Problem	08/06/2021	08/06/2021
								A070504	Premature Discharge of Battery	08/06/2021	08/06/2021
						F1903	Device Explantation	A0705	Battery Problem	08/06/2021	08/06/2021
								A070504	Premature Discharge of Battery	08/06/2021	08/06/2021
				CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem	08/06/2021	08/06/2021
								A070504	Premature Discharge of Battery	08/06/2021	08/06/2021
						F08	Hospitalization or Prolonged Hospitalization	A0705	Battery Problem	08/06/2021	08/06/2021
								A070504	Premature Discharge of Battery	08/06/2021	08/06/2021
						F1903	Device Explantation	A0705	Battery Problem	08/06/2021	08/06/2021
								A070504	Premature Discharge of Battery	08/06/2021	08/06/2021
				CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem	08/06/2021	08/06/2021
								A070504	Premature Discharge of Battery	08/06/2021	08/06/2021
						F08	Hospitalization or Prolonged Hospitalization	A0705	Battery Problem	08/06/2021	08/06/2021
								A070504	Premature Discharge of Battery	08/06/2021	08/06/2021
						F1903	Device Explantation	A0705	Battery Problem	08/06/2021	08/06/2021
								A070504	Premature Discharge of Battery	08/06/2021	08/06/2021
				CT233	Surgical	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem	08/06/2021	08/06/2021
								A070504	Premature Discharge of Battery	08/06/2021	08/06/2021
						F08	Hospitalization or Prolonged Hospitalization	A0705	Battery Problem	08/06/2021	08/06/2021
								A070504	Premature Discharge of Battery	08/06/2021	08/06/2021
						F1903	Device Explantation	A0705	Battery Problem	08/06/2021	08/06/2021
								A070504	Premature Discharge of Battery	08/06/2021	08/06/2021
				CT2406	Implantable	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem	08/06/2021	08/06/2021
								A070504	Premature Discharge of Battery	08/06/2021	08/06/2021
						F08	Hospitalization or Prolonged Hospitalization	A0705	Battery Problem	08/06/2021	08/06/2021
								A070504	Premature Discharge of Battery	08/06/2021	08/06/2021
						F1903	Device Explantation	A0705	Battery Problem	08/06/2021	08/06/2021
								A070504	Premature Discharge of Battery	08/06/2021	08/06/2021
				CT2407	Active implantable	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem	08/06/2021	08/06/2021
								A070504	Premature Discharge of Battery	08/06/2021	08/06/2021
						F08	Hospitalization or Prolonged Hospitalization	A0705	Battery Problem	08/06/2021	08/06/2021
								A070504	Premature Discharge of Battery	08/06/2021	08/06/2021
						F1903	Device Explantation	A0705	Battery Problem	08/06/2021	08/06/2021
								A070504	Premature Discharge of Battery	08/06/2021	08/06/2021
				CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem	08/06/2021	08/06/2021
								A070504	Premature Discharge of Battery	08/06/2021	08/06/2021
						F08	Hospitalization or Prolonged Hospitalization	A0705	Battery Problem	08/06/2021	08/06/2021
								A070504	Premature Discharge of Battery	08/06/2021	08/06/2021
						F1903	Device Explantation	A0705	Battery Problem	08/06/2021	08/06/2021
								A070504	Premature Discharge of Battery	08/06/2021	08/06/2021
				CT2727	Active implant devices/ancillaries	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem	08/06/2021	08/06/2021
								A070504	Premature Discharge of Battery	08/06/2021	08/06/2021
						F08	Hospitalization or Prolonged Hospitalization	A0705	Battery Problem	08/06/2021	08/06/2021
								A070504	Premature Discharge of Battery	08/06/2021	08/06/2021
						F1903	Device Explantation	A0705	Battery Problem	08/06/2021	08/06/2021
								A070504	Premature Discharge of Battery	08/06/2021	08/06/2021
				CT2750	Implantable cardiac electrical stimulators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem	08/06/2021	08/06/2021
								A070504	Premature Discharge of Battery	08/06/2021	08/06/2021
						F08	Hospitalization or Prolonged Hospitalization	A0705	Battery Problem	08/06/2021	08/06/2021
								A070504	Premature Discharge of Battery	08/06/2021	08/06/2021
						F1903	Device Explantation	A0705	Battery Problem	08/06/2021	08/06/2021
								A070504	Premature Discharge of Battery	08/06/2021	08/06/2021
				CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem	08/06/2021	08/06/2021
								A070504	Premature Discharge of Battery	08/06/2021	08/06/2021
						F08	Hospitalization or Prolonged Hospitalization	A0705	Battery Problem	08/06/2021	08/06/2021
								A070504	Premature Discharge of Battery	08/06/2021	08/06/2021
						F1903	Device Explantation	A0705	Battery Problem	08/06/2021	08/06/2021
								A070504	Premature Discharge of Battery	08/06/2021	08/06/2021
				CT2985	Active	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem	08/06/2021	08/06/2021
								A070504	Premature Discharge of Battery	08/06/2021	08/06/2021
						F08	Hospitalization or Prolonged Hospitalization	A0705	Battery Problem	08/06/2021	08/06/2021
								A070504	Premature Discharge of Battery	08/06/2021	08/06/2021
						F1903	Device Explantation	A0705	Battery Problem	08/06/2021	08/06/2021
								A070504	Premature Discharge of Battery	08/06/2021	08/06/2021
				CT3001	Implantable metal	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem	08/06/2021	08/06/2021
								A070504	Premature Discharge of Battery	08/06/2021	08/06/2021
						F08	Hospitalization or Prolonged Hospitalization	A0705	Battery Problem	08/06/2021	08/06/2021
								A070504	Premature Discharge of Battery	08/06/2021	08/06/2021
						F1903	Device Explantation	A0705	Battery Problem	08/06/2021	08/06/2021
								A070504	Premature Discharge of Battery	08/06/2021	08/06/2021
				CT3002	Implantable synthetic polymer	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem	08/06/2021	08/06/2021
								A070504	Premature Discharge of Battery	08/06/2021	08/06/2021
						F08	Hospitalization or Prolonged Hospitalization	A0705	Battery Problem	08/06/2021	08/06/2021
								A070504	Premature Discharge of Battery	08/06/2021	08/06/2021
						F1903	Device Explantation	A0705	Battery Problem	08/06/2021	08/06/2021
								A070504	Premature Discharge of Battery	08/06/2021	08/06/2021
				CT3006	Implantable inorganic material	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem	08/06/2021	08/06/2021
								A070504	Premature Discharge of Battery	08/06/2021	08/06/2021
						F08	Hospitalization or Prolonged Hospitalization	A0705	Battery Problem	08/06/2021	08/06/2021
								A070504	Premature Discharge of Battery	08/06/2021	08/06/2021
						F1903	Device Explantation	A0705	Battery Problem	08/06/2021	08/06/2021
								A070504	Premature Discharge of Battery	08/06/2021	08/06/2021

					CT314	Electrically powered	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem	08/06/2021	08/06/2021
									A070504	Premature Discharge of Battery	08/06/2021	08/06/2021
							F08	Hospitalization or Prolonged Hospitalization	A0705	Battery Problem	08/06/2021	08/06/2021
									A070504	Premature Discharge of Battery	08/06/2021	08/06/2021
							F1903	Device Explantation	A0705	Battery Problem	08/06/2021	08/06/2021
									A070504	Premature Discharge of Battery	08/06/2021	08/06/2021
					CT321	Long-term surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem	08/06/2021	08/06/2021
									A070504	Premature Discharge of Battery	08/06/2021	08/06/2021
							F08	Hospitalization or Prolonged Hospitalization	A0705	Battery Problem	08/06/2021	08/06/2021
									A070504	Premature Discharge of Battery	08/06/2021	08/06/2021
							F1903	Device Explantation	A0705	Battery Problem	08/06/2021	08/06/2021
									A070504	Premature Discharge of Battery	08/06/2021	08/06/2021
					CT449	Generators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem	08/06/2021	08/06/2021
									A070504	Premature Discharge of Battery	08/06/2021	08/06/2021
							F08	Hospitalization or Prolonged Hospitalization	A0705	Battery Problem	08/06/2021	08/06/2021
									A070504	Premature Discharge of Battery	08/06/2021	08/06/2021
							F1903	Device Explantation	A0705	Battery Problem	08/06/2021	08/06/2021
									A070504	Premature Discharge of Battery	08/06/2021	08/06/2021
					CT981	Single use	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem	08/06/2021	08/06/2021
									A070504	Premature Discharge of Battery	08/06/2021	08/06/2021
							F08	Hospitalization or Prolonged Hospitalization	A0705	Battery Problem	08/06/2021	08/06/2021
									A070504	Premature Discharge of Battery	08/06/2021	08/06/2021
							F1903	Device Explantation	A0705	Battery Problem	08/06/2021	08/06/2021
									A070504	Premature Discharge of Battery	08/06/2021	08/06/2021
					CT983	Surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem	08/06/2021	08/06/2021
									A070504	Premature Discharge of Battery	08/06/2021	08/06/2021
							F08	Hospitalization or Prolonged Hospitalization	A0705	Battery Problem	08/06/2021	08/06/2021
									A070504	Premature Discharge of Battery	08/06/2021	08/06/2021
							F1903	Device Explantation	A0705	Battery Problem	08/06/2021	08/06/2021
									A070504	Premature Discharge of Battery	08/06/2021	08/06/2021
10972	2021	All other reportable incidents	47263	Cardiac resynchronization therapy implantable pacemaker	CT1684	Cardiac pacemakers and associated devices	E011903	Syncope/Fainting	A070909	Oversensing	19/04/2021	19/04/2021
									A0712	Pacing Problem	19/04/2021	19/04/2021
									A0908	Incorrect, Inadequate or Imprecise Result or Readings	19/04/2021	19/04/2021
							E2403	No Clinical Signs, Symptoms or Conditions	A070909	Oversensing	19/04/2021	19/04/2021
									A0712	Pacing Problem	19/04/2021	19/04/2021
									A0908	Incorrect, Inadequate or Imprecise Result or Readings	19/04/2021	19/04/2021
							F08	Hospitalization or Prolonged Hospitalization	A070909	Oversensing	19/04/2021	19/04/2021
									A0712	Pacing Problem	19/04/2021	19/04/2021
									A0908	Incorrect, Inadequate or Imprecise Result or Readings	19/04/2021	19/04/2021
							F12	Serious Injury/ Illness/ Impairment	A070909	Oversensing	19/04/2021	19/04/2021
									A0712	Pacing Problem	19/04/2021	19/04/2021
									A0908	Incorrect, Inadequate or Imprecise Result or Readings	19/04/2021	19/04/2021
							F19	Surgical Intervention	A070909	Oversensing	19/04/2021	19/04/2021
									A0712	Pacing Problem	19/04/2021	19/04/2021
									A0908	Incorrect, Inadequate or Imprecise Result or Readings	19/04/2021	19/04/2021
					CT1685	Cardiac pacemakers	E011903	Syncope/Fainting	A070909	Oversensing	19/04/2021	19/04/2021
									A0712	Pacing Problem	19/04/2021	19/04/2021
									A0908	Incorrect, Inadequate or Imprecise Result or Readings	19/04/2021	19/04/2021
							E2403	No Clinical Signs, Symptoms or Conditions	A070909	Oversensing	19/04/2021	19/04/2021
									A0712	Pacing Problem	19/04/2021	19/04/2021
									A0908	Incorrect, Inadequate or Imprecise Result or Readings	19/04/2021	19/04/2021
							F08	Hospitalization or Prolonged Hospitalization	A070909	Oversensing	19/04/2021	19/04/2021
									A0712	Pacing Problem	19/04/2021	19/04/2021
									A0908	Incorrect, Inadequate or Imprecise Result or Readings	19/04/2021	19/04/2021
							F12	Serious Injury/ Illness/ Impairment	A070909	Oversensing	19/04/2021	19/04/2021
									A0712	Pacing Problem	19/04/2021	19/04/2021
									A0908	Incorrect, Inadequate or Imprecise Result or Readings	19/04/2021	19/04/2021
							F19	Surgical Intervention	A070909	Oversensing	19/04/2021	19/04/2021
									A0712	Pacing Problem	19/04/2021	19/04/2021
									A0908	Incorrect, Inadequate or Imprecise Result or Readings	19/04/2021	19/04/2021
					CT257	Implantable cardiac pacemakers	E011903	Syncope/Fainting	A070909	Oversensing	19/04/2021	19/04/2021
									A0712	Pacing Problem	19/04/2021	19/04/2021
									A0908	Incorrect, Inadequate or Imprecise Result or Readings	19/04/2021	19/04/2021
							E2403	No Clinical Signs, Symptoms or Conditions	A070909	Oversensing	19/04/2021	19/04/2021
									A0712	Pacing Problem	19/04/2021	19/04/2021
									A0908	Incorrect, Inadequate or Imprecise Result or Readings	19/04/2021	19/04/2021
							F08	Hospitalization or Prolonged Hospitalization	A070909	Oversensing	19/04/2021	19/04/2021
									A0712	Pacing Problem	19/04/2021	19/04/2021
									A0908	Incorrect, Inadequate or Imprecise Result or Readings	19/04/2021	19/04/2021
							F12	Serious Injury/ Illness/ Impairment	A070909	Oversensing	19/04/2021	19/04/2021
									A0712	Pacing Problem	19/04/2021	19/04/2021
									A0908	Incorrect, Inadequate or Imprecise Result or Readings	19/04/2021	19/04/2021
							F19	Surgical Intervention	A070909	Oversensing	19/04/2021	19/04/2021
									A0712	Pacing Problem	19/04/2021	19/04/2021
									A0908	Incorrect, Inadequate or Imprecise Result or Readings	19/04/2021	19/04/2021
					CT2866	Implantable cardiac electrical stimulators	E011903	Syncope/Fainting	A070909	Oversensing	19/04/2021	19/04/2021

									A0712	Pacing Problem	19/04/2021	19/04/2021
									A0908	Incorrect, Inadequate or Imprecise Result or Readings	19/04/2021	19/04/2021
							E2403	No Clinical Signs, Symptoms or Conditions	A070909	Oversensing	19/04/2021	19/04/2021
									A0712	Pacing Problem	19/04/2021	19/04/2021
									A0908	Incorrect, Inadequate or Imprecise Result or Readings	19/04/2021	19/04/2021
							F08	Hospitalization or Prolonged Hospitalization	A070909	Oversensing	19/04/2021	19/04/2021
									A0712	Pacing Problem	19/04/2021	19/04/2021
									A0908	Incorrect, Inadequate or Imprecise Result or Readings	19/04/2021	19/04/2021
							F12	Serious Injury/ Illness/ Impairment	A070909	Oversensing	19/04/2021	19/04/2021
									A0712	Pacing Problem	19/04/2021	19/04/2021
									A0908	Incorrect, Inadequate or Imprecise Result or Readings	19/04/2021	19/04/2021
							F19	Surgical Intervention	A070909	Oversensing	19/04/2021	19/04/2021
									A0712	Pacing Problem	19/04/2021	19/04/2021
									A0908	Incorrect, Inadequate or Imprecise Result or Readings	19/04/2021	19/04/2021
10973	2021	All other reportable incidents	47265	Dual-chamber implantable pacemaker, rate-responsive	CT141	Electrodes/Leads and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem	19/05/2021	19/05/2021
									A070504	Premature Discharge of Battery	19/05/2021	19/05/2021
							F08	Hospitalization or Prolonged Hospitalization	A0705	Battery Problem	19/05/2021	19/05/2021
									A070504	Premature Discharge of Battery	19/05/2021	19/05/2021
							F1903	Device Explantation	A0705	Battery Problem	19/05/2021	19/05/2021
									A070504	Premature Discharge of Battery	19/05/2021	19/05/2021
					CT145	Cardiovascular devices	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem	19/05/2021	19/05/2021
									A070504	Premature Discharge of Battery	19/05/2021	19/05/2021
							F08	Hospitalization or Prolonged Hospitalization	A0705	Battery Problem	19/05/2021	19/05/2021
									A070504	Premature Discharge of Battery	19/05/2021	19/05/2021
							F1903	Device Explantation	A0705	Battery Problem	19/05/2021	19/05/2021
									A070504	Premature Discharge of Battery	19/05/2021	19/05/2021
					CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem	19/05/2021	19/05/2021
									A070504	Premature Discharge of Battery	19/05/2021	19/05/2021
							F08	Hospitalization or Prolonged Hospitalization	A0705	Battery Problem	19/05/2021	19/05/2021
									A070504	Premature Discharge of Battery	19/05/2021	19/05/2021
							F1903	Device Explantation	A0705	Battery Problem	19/05/2021	19/05/2021
									A070504	Premature Discharge of Battery	19/05/2021	19/05/2021
					CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem	19/05/2021	19/05/2021
									A070504	Premature Discharge of Battery	19/05/2021	19/05/2021
							F08	Hospitalization or Prolonged Hospitalization	A0705	Battery Problem	19/05/2021	19/05/2021
									A070504	Premature Discharge of Battery	19/05/2021	19/05/2021
							F1903	Device Explantation	A0705	Battery Problem	19/05/2021	19/05/2021
									A070504	Premature Discharge of Battery	19/05/2021	19/05/2021
					CT233	Surgical	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem	19/05/2021	19/05/2021
									A070504	Premature Discharge of Battery	19/05/2021	19/05/2021
							F08	Hospitalization or Prolonged Hospitalization	A0705	Battery Problem	19/05/2021	19/05/2021
									A070504	Premature Discharge of Battery	19/05/2021	19/05/2021
							F1903	Device Explantation	A0705	Battery Problem	19/05/2021	19/05/2021
									A070504	Premature Discharge of Battery	19/05/2021	19/05/2021
					CT2406	Implantable	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem	19/05/2021	19/05/2021
									A070504	Premature Discharge of Battery	19/05/2021	19/05/2021
							F08	Hospitalization or Prolonged Hospitalization	A0705	Battery Problem	19/05/2021	19/05/2021
									A070504	Premature Discharge of Battery	19/05/2021	19/05/2021
							F1903	Device Explantation	A0705	Battery Problem	19/05/2021	19/05/2021
									A070504	Premature Discharge of Battery	19/05/2021	19/05/2021
					CT2407	Active implantable	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem	19/05/2021	19/05/2021
									A070504	Premature Discharge of Battery	19/05/2021	19/05/2021
							F08	Hospitalization or Prolonged Hospitalization	A0705	Battery Problem	19/05/2021	19/05/2021
									A070504	Premature Discharge of Battery	19/05/2021	19/05/2021
							F1903	Device Explantation	A0705	Battery Problem	19/05/2021	19/05/2021
									A070504	Premature Discharge of Battery	19/05/2021	19/05/2021
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem	19/05/2021	19/05/2021
									A070504	Premature Discharge of Battery	19/05/2021	19/05/2021
							F08	Hospitalization or Prolonged Hospitalization	A0705	Battery Problem	19/05/2021	19/05/2021
									A070504	Premature Discharge of Battery	19/05/2021	19/05/2021
							F1903	Device Explantation	A0705	Battery Problem	19/05/2021	19/05/2021
									A070504	Premature Discharge of Battery	19/05/2021	19/05/2021
					CT2727	Active implant devices/ancillaries	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem	19/05/2021	19/05/2021
									A070504	Premature Discharge of Battery	19/05/2021	19/05/2021
							F08	Hospitalization or Prolonged Hospitalization	A0705	Battery Problem	19/05/2021	19/05/2021
									A070504	Premature Discharge of Battery	19/05/2021	19/05/2021
							F1903	Device Explantation	A0705	Battery Problem	19/05/2021	19/05/2021
									A070504	Premature Discharge of Battery	19/05/2021	19/05/2021
					CT2750	Implantable cardiac electrical stimulators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem	19/05/2021	19/05/2021
									A070504	Premature Discharge of Battery	19/05/2021	19/05/2021
							F08	Hospitalization or Prolonged Hospitalization	A0705	Battery Problem	19/05/2021	19/05/2021
									A070504	Premature Discharge of Battery	19/05/2021	19/05/2021
							F1903	Device Explantation	A0705	Battery Problem	19/05/2021	19/05/2021
									A070504	Premature Discharge of Battery	19/05/2021	19/05/2021
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem	19/05/2021	19/05/2021
									A070504	Premature Discharge of Battery	19/05/2021	19/05/2021
							F08	Hospitalization or Prolonged Hospitalization	A0705	Battery Problem	19/05/2021	19/05/2021
									A070504	Premature Discharge of Battery	19/05/2021	19/05/2021
							F1903	Device Explantation	A0705	Battery Problem	19/05/2021	19/05/2021
									A070504	Premature Discharge of Battery	19/05/2021	19/05/2021
					CT2985	Active	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem	19/05/2021	19/05/2021
									A070504	Premature Discharge of Battery	19/05/2021	19/05/2021
							F08	Hospitalization or Prolonged Hospitalization	A0705	Battery Problem	19/05/2021	19/05/2021
									A070504	Premature Discharge of Battery	19/05/2021	19/05/2021

							F1903	Device Explantation	A0705	Battery Problem	19/05/2021	19/05/2021
									A070504	Premature Discharge of Battery	19/05/2021	19/05/2021
					CT3001	Implantable metal	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem	19/05/2021	19/05/2021
									A070504	Premature Discharge of Battery	19/05/2021	19/05/2021
							F08	Hospitalization or Prolonged Hospitalization	A0705	Battery Problem	19/05/2021	19/05/2021
									A070504	Premature Discharge of Battery	19/05/2021	19/05/2021
							F1903	Device Explantation	A0705	Battery Problem	19/05/2021	19/05/2021
									A070504	Premature Discharge of Battery	19/05/2021	19/05/2021
					CT3002	Implantable synthetic polymer	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem	19/05/2021	19/05/2021
									A070504	Premature Discharge of Battery	19/05/2021	19/05/2021
							F08	Hospitalization or Prolonged Hospitalization	A0705	Battery Problem	19/05/2021	19/05/2021
									A070504	Premature Discharge of Battery	19/05/2021	19/05/2021
							F1903	Device Explantation	A0705	Battery Problem	19/05/2021	19/05/2021
									A070504	Premature Discharge of Battery	19/05/2021	19/05/2021
					CT3006	Implantable inorganic material	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem	19/05/2021	19/05/2021
									A070504	Premature Discharge of Battery	19/05/2021	19/05/2021
							F08	Hospitalization or Prolonged Hospitalization	A0705	Battery Problem	19/05/2021	19/05/2021
									A070504	Premature Discharge of Battery	19/05/2021	19/05/2021
							F1903	Device Explantation	A0705	Battery Problem	19/05/2021	19/05/2021
									A070504	Premature Discharge of Battery	19/05/2021	19/05/2021
					CT314	Electrically powered	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem	19/05/2021	19/05/2021
									A070504	Premature Discharge of Battery	19/05/2021	19/05/2021
							F08	Hospitalization or Prolonged Hospitalization	A0705	Battery Problem	19/05/2021	19/05/2021
									A070504	Premature Discharge of Battery	19/05/2021	19/05/2021
							F1903	Device Explantation	A0705	Battery Problem	19/05/2021	19/05/2021
									A070504	Premature Discharge of Battery	19/05/2021	19/05/2021
					CT321	Long-term surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem	19/05/2021	19/05/2021
									A070504	Premature Discharge of Battery	19/05/2021	19/05/2021
							F08	Hospitalization or Prolonged Hospitalization	A0705	Battery Problem	19/05/2021	19/05/2021
									A070504	Premature Discharge of Battery	19/05/2021	19/05/2021
							F1903	Device Explantation	A0705	Battery Problem	19/05/2021	19/05/2021
									A070504	Premature Discharge of Battery	19/05/2021	19/05/2021
					CT449	Generators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem	19/05/2021	19/05/2021
									A070504	Premature Discharge of Battery	19/05/2021	19/05/2021
							F08	Hospitalization or Prolonged Hospitalization	A0705	Battery Problem	19/05/2021	19/05/2021
									A070504	Premature Discharge of Battery	19/05/2021	19/05/2021
							F1903	Device Explantation	A0705	Battery Problem	19/05/2021	19/05/2021
									A070504	Premature Discharge of Battery	19/05/2021	19/05/2021
					CT981	Single use	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem	19/05/2021	19/05/2021
									A070504	Premature Discharge of Battery	19/05/2021	19/05/2021
							F08	Hospitalization or Prolonged Hospitalization	A0705	Battery Problem	19/05/2021	19/05/2021
									A070504	Premature Discharge of Battery	19/05/2021	19/05/2021
							F1903	Device Explantation	A0705	Battery Problem	19/05/2021	19/05/2021
									A070504	Premature Discharge of Battery	19/05/2021	19/05/2021
					CT983	Surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem	19/05/2021	19/05/2021
									A070504	Premature Discharge of Battery	19/05/2021	19/05/2021
							F08	Hospitalization or Prolonged Hospitalization	A0705	Battery Problem	19/05/2021	19/05/2021
									A070504	Premature Discharge of Battery	19/05/2021	19/05/2021
							F1903	Device Explantation	A0705	Battery Problem	19/05/2021	19/05/2021
									A070504	Premature Discharge of Battery	19/05/2021	19/05/2021
10974	2021	All other reportable incidents	47265	Dual-chamber implantable pacemaker, rate-responsive	CT141	Electrodes/Leads and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem		
							F26	No Health Consequences or Impact	A0705	Battery Problem		
					CT145	Cardiovascular devices	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem		
							F26	No Health Consequences or Impact	A0705	Battery Problem		
					CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem		
							F26	No Health Consequences or Impact	A0705	Battery Problem		
					CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem		
							F26	No Health Consequences or Impact	A0705	Battery Problem		
					CT233	Surgical	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem		
							F26	No Health Consequences or Impact	A0705	Battery Problem		
					CT2406	Implantable	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem		
							F26	No Health Consequences or Impact	A0705	Battery Problem		
					CT2407	Active implantable	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem		
							F26	No Health Consequences or Impact	A0705	Battery Problem		
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem		
							F26	No Health Consequences or Impact	A0705	Battery Problem		
					CT2727	Active implant devices/ancillaries	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem		
							F26	No Health Consequences or Impact	A0705	Battery Problem		
					CT2750	Implantable cardiac electrical stimulators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem		
							F26	No Health Consequences or Impact	A0705	Battery Problem		
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem		
							F26	No Health Consequences or Impact	A0705	Battery Problem		
					CT2985	Active	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem		
							F26	No Health Consequences or Impact	A0705	Battery Problem		
					CT3001	Implantable metal	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem		
							F26	No Health Consequences or Impact	A0705	Battery Problem		
					CT3002	Implantable synthetic polymer	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem		
							F26	No Health Consequences or Impact	A0705	Battery Problem		
					CT3006	Implantable inorganic material	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem		
							F26	No Health Consequences or Impact	A0705	Battery Problem		
					CT314	Electrically powered	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem		
							F26	No Health Consequences or Impact	A0705	Battery Problem		
					CT321	Long-term surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem		
							F26	No Health Consequences or Impact	A0705	Battery Problem		
					CT449	Generators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem		
							F26	No Health Consequences or Impact	A0705	Battery Problem		
					CT981	Single use	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem		
							F26	No Health Consequences or Impact	A0705	Battery Problem		

					CT983	Surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem		
							F26	No Health Consequences or Impact	A0705	Battery Problem		
10975	2021	All other reportable incidents	47263	Cardiac resynchronization therapy implantable pacemaker	CT1684	Cardiac pacemakers and associated devices	E0112	Dizziness	A070101	Failure to Capture		
							E0905	Vertigo	A070101	Failure to Capture		
							F24	Insufficient Health Impact Information	A070101	Failure to Capture		
					CT1685	Cardiac pacemakers	E0112	Dizziness	A070101	Failure to Capture		
							E0905	Vertigo	A070101	Failure to Capture		
							F24	Insufficient Health Impact Information	A070101	Failure to Capture		
					CT257	Implantable cardiac pacemakers	E0112	Dizziness	A070101	Failure to Capture		
							E0905	Vertigo	A070101	Failure to Capture		
							F24	Insufficient Health Impact Information	A070101	Failure to Capture		
					CT2866	Implantable cardiac electrical stimulators	E0112	Dizziness	A070101	Failure to Capture		
							E0905	Vertigo	A070101	Failure to Capture		
							F24	Insufficient Health Impact Information	A070101	Failure to Capture		
10976	2021	All other reportable incidents	35852	Single-chamber implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E060101	Asystole	A070909	Oversensing		
									A0712	Pacing Problem		
							E0602	Cardiac Arrest	A070909	Oversensing		
									A0712	Pacing Problem		
							F26	No Health Consequences or Impact	A070909	Oversensing		
									A0712	Pacing Problem		
10977	2021	All other reportable incidents	36241	Pacemaker/defibrillator lead	CT1684	Cardiac pacemakers and associated devices	E0623	Unspecified Heart Problem	A051201	Device Dislodged or Dislocated		
									A070102	High Capture Threshold		
							E2114	Perforation	A051201	Device Dislodged or Dislocated		
									A070102	High Capture Threshold		
							E2330	Pain	A051201	Device Dislodged or Dislocated		
									A070102	High Capture Threshold		
							E233001	Chest Pain	A051201	Device Dislodged or Dislocated		
									A070102	High Capture Threshold		
							F08	Hospitalization or Prolonged Hospitalization	A051201	Device Dislodged or Dislocated		
									A070102	High Capture Threshold		
							F12	Serious Injury/ Illness/ Impairment	A051201	Device Dislodged or Dislocated		
									A070102	High Capture Threshold		
10978	2021	All other reportable incidents	47265	Dual-chamber implantable pacemaker, rate-responsive	CT141	Electrodes/Leads and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
					CT145	Cardiovascular devices	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
					CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
					CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
					CT233	Surgical	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
					CT2406	Implantable	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
					CT2407	Active implantable	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
					CT2727	Active implant devices/ancillaries	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
					CT2750	Implantable cardiac electrical stimulators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
					CT2985	Active	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
					CT3001	Implantable metal	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
					CT3002	Implantable synthetic polymer	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		

					CT3006	Implantable inorganic material	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
					CT314	Electrically powered	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
					CT321	Long-term surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
					CT449	Generators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
					CT981	Single use	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
					CT983	Surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
10979	2021	All other reportable incidents	17882	Physiologic-monitoring defibrillation system	CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A07	Electrical /Electronic Property Problem		
							F26	No Health Consequences or Impact	A07	Electrical /Electronic Property Problem		
10980	2021	All other reportable incidents	47267	Single-chamber implantable pacemaker, rate-responsive	CT141	Electrodes/Leads and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT145	Cardiovascular devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT233	Surgical	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT2406	Implantable	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT2407	Active implantable	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT2727	Active implant devices/ancillaries	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT2750	Implantable cardiac electrical stimulators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT2985	Active	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT3001	Implantable metal	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT3002	Implantable synthetic polymer	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT3006	Implantable inorganic material	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT314	Electrically powered	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT321	Long-term surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT449	Generators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT981	Single use	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT983	Surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
10981	2021	All other reportable incidents	35852	Single-chamber implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A160601	Premature Elective Replacement Indicator	23/03/2021	23/03/2021
							F1905	Device Revision or Replacement	A160601	Premature Elective Replacement Indicator	23/03/2021	23/03/2021
10982	2021	All other reportable incidents	47270	Cardiac resynchronization therapy implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	23/04/2021	23/04/2021
							F1901	Additional Surgery	A070504	Premature Discharge of Battery	23/04/2021	23/04/2021
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	23/04/2021	23/04/2021
10983	2021	All other reportable incidents	47265	Dual-chamber implantable pacemaker, rate-responsive	CT141	Electrodes/Leads and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	16/04/2021	16/04/2021
							F1901	Additional Surgery	A070504	Premature Discharge of Battery	16/04/2021	16/04/2021
					CT145	Cardiovascular devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	16/04/2021	16/04/2021
							F1901	Additional Surgery	A070504	Premature Discharge of Battery	16/04/2021	16/04/2021
					CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	16/04/2021	16/04/2021
							F1901	Additional Surgery	A070504	Premature Discharge of Battery	16/04/2021	16/04/2021
					CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	16/04/2021	16/04/2021
							F1901	Additional Surgery	A070504	Premature Discharge of Battery	16/04/2021	16/04/2021
					CT233	Surgical	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	16/04/2021	16/04/2021
							F1901	Additional Surgery	A070504	Premature Discharge of Battery	16/04/2021	16/04/2021
					CT2406	Implantable	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	16/04/2021	16/04/2021
							F1901	Additional Surgery	A070504	Premature Discharge of Battery	16/04/2021	16/04/2021
					CT2407	Active implantable	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	16/04/2021	16/04/2021
							F1901	Additional Surgery	A070504	Premature Discharge of Battery	16/04/2021	16/04/2021
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	16/04/2021	16/04/2021
							F1901	Additional Surgery	A070504	Premature Discharge of Battery	16/04/2021	16/04/2021
					CT2727	Active implant devices/ancillaries	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	16/04/2021	16/04/2021

							F1901	Additional Surgery	A070504	Premature Discharge of Battery	16/04/2021	16/04/2021
					CT2750	Implantable cardiac electrical stimulators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	16/04/2021	16/04/2021
							F1901	Additional Surgery	A070504	Premature Discharge of Battery	16/04/2021	16/04/2021
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	16/04/2021	16/04/2021
							F1901	Additional Surgery	A070504	Premature Discharge of Battery	16/04/2021	16/04/2021
					CT2985	Active	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	16/04/2021	16/04/2021
							F1901	Additional Surgery	A070504	Premature Discharge of Battery	16/04/2021	16/04/2021
					CT3001	Implantable metal	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	16/04/2021	16/04/2021
							F1901	Additional Surgery	A070504	Premature Discharge of Battery	16/04/2021	16/04/2021
					CT3002	Implantable synthetic polymer	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	16/04/2021	16/04/2021
							F1901	Additional Surgery	A070504	Premature Discharge of Battery	16/04/2021	16/04/2021
					CT3006	Implantable inorganic material	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	16/04/2021	16/04/2021
							F1901	Additional Surgery	A070504	Premature Discharge of Battery	16/04/2021	16/04/2021
					CT314	Electrically powered	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	16/04/2021	16/04/2021
							F1901	Additional Surgery	A070504	Premature Discharge of Battery	16/04/2021	16/04/2021
					CT321	Long-term surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	16/04/2021	16/04/2021
							F1901	Additional Surgery	A070504	Premature Discharge of Battery	16/04/2021	16/04/2021
					CT449	Generators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	16/04/2021	16/04/2021
							F1901	Additional Surgery	A070504	Premature Discharge of Battery	16/04/2021	16/04/2021
					CT981	Single use	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	16/04/2021	16/04/2021
							F1901	Additional Surgery	A070504	Premature Discharge of Battery	16/04/2021	16/04/2021
					CT983	Surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	16/04/2021	16/04/2021
							F1901	Additional Surgery	A070504	Premature Discharge of Battery	16/04/2021	16/04/2021
10984	2021	All other reportable incidents	35223	Endocardial/interventricular septal pacing lead	CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A0701	Capturing Problem	22/04/2021	22/04/2021
									A070911	Sensing Intermittently	22/04/2021	22/04/2021
							F1908	Prolonged Surgery	A0701	Capturing Problem	22/04/2021	22/04/2021
									A070911	Sensing Intermittently	22/04/2021	22/04/2021
10985	2021	All other reportable incidents	37265	Dual-chamber implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem	07/06/2021	07/06/2021
									A070504	Premature Discharge of Battery	07/06/2021	07/06/2021
							F08	Hospitalization or Prolonged Hospitalization	A0705	Battery Problem	07/06/2021	07/06/2021
									A070504	Premature Discharge of Battery	07/06/2021	07/06/2021
							F1905	Device Revision or Replacement	A0705	Battery Problem	07/06/2021	07/06/2021
									A070504	Premature Discharge of Battery	07/06/2021	07/06/2021
10986	2021	All other reportable incidents	35852	Single-chamber implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem	04/06/2021	04/06/2021
									A070504	Premature Discharge of Battery	04/06/2021	04/06/2021
							F08	Hospitalization or Prolonged Hospitalization	A0705	Battery Problem	04/06/2021	04/06/2021
									A070504	Premature Discharge of Battery	04/06/2021	04/06/2021
							F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A0705	Battery Problem	04/06/2021	04/06/2021
									A070504	Premature Discharge of Battery	04/06/2021	04/06/2021
							F1905	Device Revision or Replacement	A0705	Battery Problem	04/06/2021	04/06/2021
									A070504	Premature Discharge of Battery	04/06/2021	04/06/2021
10987	2021	All other reportable incidents	35852	Single-chamber implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	01/06/2021	01/06/2021
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	01/06/2021	01/06/2021
							F19	Surgical Intervention	A070504	Premature Discharge of Battery	01/06/2021	01/06/2021
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	01/06/2021	01/06/2021
10988	2021	All other reportable incidents	35852	Single-chamber implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	19/04/2021	19/04/2021
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	19/04/2021	19/04/2021
							F19	Surgical Intervention	A070504	Premature Discharge of Battery	19/04/2021	19/04/2021
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	19/04/2021	19/04/2021
10989	2021	All other reportable incidents	47265	Dual-chamber implantable pacemaker, rate-responsive	CT141	Electrodes/Leads and associated devices	E060101	Asystole	A070909	Oversensing		
									A0712	Pacing Problem		
									A090801	Signal Artifact/Noise		
							E2403	No Clinical Signs, Symptoms or Conditions	A070909	Oversensing		
									A0712	Pacing Problem		
									A090801	Signal Artifact/Noise		
							F26	No Health Consequences or Impact	A070909	Oversensing		
									A0712	Pacing Problem		
									A090801	Signal Artifact/Noise		
					CT145	Cardiovascular devices	E060101	Asystole	A070909	Oversensing		
									A0712	Pacing Problem		
									A090801	Signal Artifact/Noise		
							E2403	No Clinical Signs, Symptoms or Conditions	A070909	Oversensing		
									A0712	Pacing Problem		
									A090801	Signal Artifact/Noise		
							F26	No Health Consequences or Impact	A070909	Oversensing		
									A0712	Pacing Problem		
									A090801	Signal Artifact/Noise		
					CT1684	Cardiac pacemakers and associated devices	E060101	Asystole	A070909	Oversensing		
									A0712	Pacing Problem		
									A090801	Signal Artifact/Noise		
							E2403	No Clinical Signs, Symptoms or Conditions	A070909	Oversensing		
									A0712	Pacing Problem		
									A090801	Signal Artifact/Noise		
							F26	No Health Consequences or Impact	A070909	Oversensing		
									A0712	Pacing Problem		
									A090801	Signal Artifact/Noise		
					CT1685	Cardiac pacemakers	E060101	Asystole	A070909	Oversensing		
									A0712	Pacing Problem		
									A090801	Signal Artifact/Noise		
							E2403	No Clinical Signs, Symptoms or Conditions	A070909	Oversensing		
									A0712	Pacing Problem		
									A090801	Signal Artifact/Noise		
							F26	No Health Consequences or Impact	A070909	Oversensing		
									A0712	Pacing Problem		
									A090801	Signal Artifact/Noise		
					CT233	Surgical	E060101	Asystole	A070909	Oversensing		
									A0712	Pacing Problem		
									A090801	Signal Artifact/Noise		

							E2403	No Clinical Signs, Symptoms or Conditions	A070909	Oversensing		
									A0712	Pacing Problem		
									A090801	Signal Artifact/Noise		
							F26	No Health Consequences or Impact	A070909	Oversensing		
									A0712	Pacing Problem		
									A090801	Signal Artifact/Noise		
					CT2406	Implantable	E060101	Asystole	A070909	Oversensing		
									A0712	Pacing Problem		
									A090801	Signal Artifact/Noise		
							E2403	No Clinical Signs, Symptoms or Conditions	A070909	Oversensing		
									A0712	Pacing Problem		
									A090801	Signal Artifact/Noise		
							F26	No Health Consequences or Impact	A070909	Oversensing		
									A0712	Pacing Problem		
									A090801	Signal Artifact/Noise		
					CT2407	Active implantable	E060101	Asystole	A070909	Oversensing		
									A0712	Pacing Problem		
									A090801	Signal Artifact/Noise		
							E2403	No Clinical Signs, Symptoms or Conditions	A070909	Oversensing		
									A0712	Pacing Problem		
									A090801	Signal Artifact/Noise		
							F26	No Health Consequences or Impact	A070909	Oversensing		
									A0712	Pacing Problem		
									A090801	Signal Artifact/Noise		
					CT257	Implantable cardiac pacemakers	E060101	Asystole	A070909	Oversensing		
									A0712	Pacing Problem		
									A090801	Signal Artifact/Noise		
							E2403	No Clinical Signs, Symptoms or Conditions	A070909	Oversensing		
									A0712	Pacing Problem		
									A090801	Signal Artifact/Noise		
							F26	No Health Consequences or Impact	A070909	Oversensing		
									A0712	Pacing Problem		
									A090801	Signal Artifact/Noise		
					CT2727	Active implant devices/ancillaries	E060101	Asystole	A070909	Oversensing		
									A0712	Pacing Problem		
									A090801	Signal Artifact/Noise		
							E2403	No Clinical Signs, Symptoms or Conditions	A070909	Oversensing		
									A0712	Pacing Problem		
									A090801	Signal Artifact/Noise		
							F26	No Health Consequences or Impact	A070909	Oversensing		
									A0712	Pacing Problem		
									A090801	Signal Artifact/Noise		
					CT2750	Implantable cardiac electrical stimulators and associated devices	E060101	Asystole	A070909	Oversensing		
									A0712	Pacing Problem		
									A090801	Signal Artifact/Noise		
							E2403	No Clinical Signs, Symptoms or Conditions	A070909	Oversensing		
									A0712	Pacing Problem		
									A090801	Signal Artifact/Noise		
							F26	No Health Consequences or Impact	A070909	Oversensing		
									A0712	Pacing Problem		
									A090801	Signal Artifact/Noise		
					CT2866	Implantable cardiac electrical stimulators	E060101	Asystole	A070909	Oversensing		
									A0712	Pacing Problem		
									A090801	Signal Artifact/Noise		
							E2403	No Clinical Signs, Symptoms or Conditions	A070909	Oversensing		
									A0712	Pacing Problem		
									A090801	Signal Artifact/Noise		
							F26	No Health Consequences or Impact	A070909	Oversensing		
									A0712	Pacing Problem		
									A090801	Signal Artifact/Noise		
					CT2985	Active	E060101	Asystole	A070909	Oversensing		
									A0712	Pacing Problem		
									A090801	Signal Artifact/Noise		
							E2403	No Clinical Signs, Symptoms or Conditions	A070909	Oversensing		
									A0712	Pacing Problem		
									A090801	Signal Artifact/Noise		
							F26	No Health Consequences or Impact	A070909	Oversensing		
									A0712	Pacing Problem		
									A090801	Signal Artifact/Noise		
					CT3001	Implantable metal	E060101	Asystole	A070909	Oversensing		
									A0712	Pacing Problem		
									A090801	Signal Artifact/Noise		
							E2403	No Clinical Signs, Symptoms or Conditions	A070909	Oversensing		
									A0712	Pacing Problem		
									A090801	Signal Artifact/Noise		
							F26	No Health Consequences or Impact	A070909	Oversensing		
									A0712	Pacing Problem		
									A090801	Signal Artifact/Noise		
					CT3002	Implantable synthetic polymer	E060101	Asystole	A070909	Oversensing		
									A0712	Pacing Problem		
									A090801	Signal Artifact/Noise		
							E2403	No Clinical Signs, Symptoms or Conditions	A070909	Oversensing		
									A0712	Pacing Problem		
									A090801	Signal Artifact/Noise		
							F26	No Health Consequences or Impact	A070909	Oversensing		
									A0712	Pacing Problem		
									A090801	Signal Artifact/Noise		
					CT3006	Implantable inorganic material	E060101	Asystole	A070909	Oversensing		

									A0712	Pacing Problem		
									A090801	Signal Artifact/Noise		
							E2403	No Clinical Signs, Symptoms or Conditions	A070909	Oversensing		
									A0712	Pacing Problem		
									A090801	Signal Artifact/Noise		
							F26	No Health Consequences or Impact	A070909	Oversensing		
									A0712	Pacing Problem		
									A090801	Signal Artifact/Noise		
					CT314	Electrically powered	E060101	Asystole	A070909	Oversensing		
									A0712	Pacing Problem		
									A090801	Signal Artifact/Noise		
							E2403	No Clinical Signs, Symptoms or Conditions	A070909	Oversensing		
									A0712	Pacing Problem		
									A090801	Signal Artifact/Noise		
							F26	No Health Consequences or Impact	A070909	Oversensing		
									A0712	Pacing Problem		
									A090801	Signal Artifact/Noise		
					CT321	Long-term surgical invasive	E060101	Asystole	A070909	Oversensing		
									A0712	Pacing Problem		
									A090801	Signal Artifact/Noise		
							E2403	No Clinical Signs, Symptoms or Conditions	A070909	Oversensing		
									A0712	Pacing Problem		
									A090801	Signal Artifact/Noise		
							F26	No Health Consequences or Impact	A070909	Oversensing		
									A0712	Pacing Problem		
									A090801	Signal Artifact/Noise		
					CT449	Generators and associated devices	E060101	Asystole	A070909	Oversensing		
									A0712	Pacing Problem		
									A090801	Signal Artifact/Noise		
							E2403	No Clinical Signs, Symptoms or Conditions	A070909	Oversensing		
									A0712	Pacing Problem		
									A090801	Signal Artifact/Noise		
							F26	No Health Consequences or Impact	A070909	Oversensing		
									A0712	Pacing Problem		
									A090801	Signal Artifact/Noise		
					CT981	Single use	E060101	Asystole	A070909	Oversensing		
									A0712	Pacing Problem		
									A090801	Signal Artifact/Noise		
							E2403	No Clinical Signs, Symptoms or Conditions	A070909	Oversensing		
									A0712	Pacing Problem		
									A090801	Signal Artifact/Noise		
							F26	No Health Consequences or Impact	A070909	Oversensing		
									A0712	Pacing Problem		
									A090801	Signal Artifact/Noise		
					CT983	Surgical invasive	E060101	Asystole	A070909	Oversensing		
									A0712	Pacing Problem		
									A090801	Signal Artifact/Noise		
							E2403	No Clinical Signs, Symptoms or Conditions	A070909	Oversensing		
									A0712	Pacing Problem		
									A090801	Signal Artifact/Noise		
							F26	No Health Consequences or Impact	A070909	Oversensing		
									A0712	Pacing Problem		
									A090801	Signal Artifact/Noise		
10990	2021	All other reportable incidents	47265	Dual-chamber implantable pacemaker, rate-responsive	CT141	Electrodes/Leads and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem		
							F26	No Health Consequences or Impact	A0705	Battery Problem		
					CT145	Cardiovascular devices	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem		
							F26	No Health Consequences or Impact	A0705	Battery Problem		
					CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem		
							F26	No Health Consequences or Impact	A0705	Battery Problem		
					CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem		
							F26	No Health Consequences or Impact	A0705	Battery Problem		
					CT233	Surgical	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem		
							F26	No Health Consequences or Impact	A0705	Battery Problem		
					CT2406	Implantable	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem		
							F26	No Health Consequences or Impact	A0705	Battery Problem		
					CT2407	Active implantable	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem		
							F26	No Health Consequences or Impact	A0705	Battery Problem		
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem		
							F26	No Health Consequences or Impact	A0705	Battery Problem		
					CT2727	Active implant devices/ancillaries	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem		
							F26	No Health Consequences or Impact	A0705	Battery Problem		
					CT2750	Implantable cardiac electrical stimulators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem		
							F26	No Health Consequences or Impact	A0705	Battery Problem		
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem		
							F26	No Health Consequences or Impact	A0705	Battery Problem		
					CT2985	Active	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem		
							F26	No Health Consequences or Impact	A0705	Battery Problem		
					CT3001	Implantable metal	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem		
							F26	No Health Consequences or Impact	A0705	Battery Problem		
					CT3002	Implantable synthetic polymer	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem		
							F26	No Health Consequences or Impact	A0705	Battery Problem		
					CT3006	Implantable inorganic material	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem		
							F26	No Health Consequences or Impact	A0705	Battery Problem		
					CT314	Electrically powered	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem		
							F26	No Health Consequences or Impact	A0705	Battery Problem		
					CT321	Long-term surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem		
							F26	No Health Consequences or Impact	A0705	Battery Problem		
					CT449	Generators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem		

						F26	No Health Consequences or Impact	A0705	Battery Problem			
				CT981	Single use	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem			
						F26	No Health Consequences or Impact	A0705	Battery Problem			
				CT983	Surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem			
						F26	No Health Consequences or Impact	A0705	Battery Problem			
10991	2021	All other reportable incidents	47265	Dual-chamber implantable pacemaker, rate-responsive	CT141	Electrodes/Leads and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem		
							F26	No Health Consequences or Impact	A0705	Battery Problem		
					CT145	Cardiovascular devices	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem		
							F26	No Health Consequences or Impact	A0705	Battery Problem		
					CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem		
							F26	No Health Consequences or Impact	A0705	Battery Problem		
					CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem		
							F26	No Health Consequences or Impact	A0705	Battery Problem		
					CT233	Surgical	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem		
							F26	No Health Consequences or Impact	A0705	Battery Problem		
					CT2406	Implantable	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem		
							F26	No Health Consequences or Impact	A0705	Battery Problem		
					CT2407	Active implantable	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem		
							F26	No Health Consequences or Impact	A0705	Battery Problem		
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem		
							F26	No Health Consequences or Impact	A0705	Battery Problem		
					CT2727	Active implant devices/ancillaries	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem		
							F26	No Health Consequences or Impact	A0705	Battery Problem		
					CT2750	Implantable cardiac electrical stimulators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem		
							F26	No Health Consequences or Impact	A0705	Battery Problem		
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem		
							F26	No Health Consequences or Impact	A0705	Battery Problem		
					CT2985	Active	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem		
							F26	No Health Consequences or Impact	A0705	Battery Problem		
					CT3001	Implantable metal	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem		
							F26	No Health Consequences or Impact	A0705	Battery Problem		
					CT3002	Implantable synthetic polymer	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem		
							F26	No Health Consequences or Impact	A0705	Battery Problem		
					CT3006	Implantable inorganic material	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem		
							F26	No Health Consequences or Impact	A0705	Battery Problem		
					CT314	Electrically powered	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem		
							F26	No Health Consequences or Impact	A0705	Battery Problem		
					CT321	Long-term surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem		
							F26	No Health Consequences or Impact	A0705	Battery Problem		
					CT449	Generators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem		
							F26	No Health Consequences or Impact	A0705	Battery Problem		
					CT981	Single use	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem		
							F26	No Health Consequences or Impact	A0705	Battery Problem		
					CT983	Surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem		
							F26	No Health Consequences or Impact	A0705	Battery Problem		
10992	2021	All other reportable incidents	35223	Endocardial/interventricular septal pacing lead	CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A090407	Therapy Delivered to Incorrect Body Area		
							F1901	Additional Surgery	A090407	Therapy Delivered to Incorrect Body Area		
							F1905	Device Revision or Replacement	A090407	Therapy Delivered to Incorrect Body Area		
10993	2021	All other reportable incidents	47270	Cardiac resynchronization therapy implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
10994	2021	All other reportable incidents	35852	Single-chamber implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem	04/08/2021	04/08/2021
									A070504	Premature Discharge of Battery	04/08/2021	04/08/2021
							F08	Hospitalization or Prolonged Hospitalization	A0705	Battery Problem	04/08/2021	04/08/2021
									A070504	Premature Discharge of Battery	04/08/2021	04/08/2021
							F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A0705	Battery Problem	04/08/2021	04/08/2021
									A070504	Premature Discharge of Battery	04/08/2021	04/08/2021
							F1905	Device Revision or Replacement	A0705	Battery Problem	04/08/2021	04/08/2021
									A070504	Premature Discharge of Battery	04/08/2021	04/08/2021
10995	2021	All other reportable incidents	47265	Dual-chamber implantable pacemaker, rate-responsive	CT141	Electrodes/Leads and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT145	Cardiovascular devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT233	Surgical	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT2406	Implantable	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT2407	Active implantable	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT2727	Active implant devices/ancillaries	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT2750	Implantable cardiac electrical stimulators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT2985	Active	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT3001	Implantable metal	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT3002	Implantable synthetic polymer	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT3006	Implantable inorganic material	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		

					CT314	Electrically powered	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT321	Long-term surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT449	Generators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT981	Single use	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT983	Surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
10996	2021	All other reportable incidents	35223	Endocardial/interventricular septal pacing lead	CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A0701	Capturing Problem	22/04/2021	22/04/2021
									A0709	Device Sensing Problem	22/04/2021	22/04/2021
							F1905	Device Revision or Replacement	A0701	Capturing Problem	22/04/2021	22/04/2021
									A0709	Device Sensing Problem	22/04/2021	22/04/2021
							F1908	Prolonged Surgery	A0701	Capturing Problem	22/04/2021	22/04/2021
									A0709	Device Sensing Problem	22/04/2021	22/04/2021
10997	2021	All other reportable incidents	35223	Endocardial/interventricular septal pacing lead	CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070909	Oversensing		
									A0712	Pacing Problem		
							F1901	Additional Surgery	A070909	Oversensing		
									A0712	Pacing Problem		
							F1905	Device Revision or Replacement	A070909	Oversensing		
									A0712	Pacing Problem		
10998	2021	All other reportable incidents	36241	Pacemaker/defibrillator lead	CT1684	Cardiac pacemakers and associated devices	E060101	Asystole	A0712	Pacing Problem		
							E0602	Cardiac Arrest	A0712	Pacing Problem		
							F05	Delay to Treatment/ Therapy	A0712	Pacing Problem		
							F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A0712	Pacing Problem		
							F19	Surgical Intervention	A0712	Pacing Problem		
10999	2021	All other reportable incidents	47265	Dual-chamber implantable pacemaker, rate-responsive	CT141	Electrodes/Leads and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	02/07/2021	02/07/2021
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	02/07/2021	02/07/2021
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	02/07/2021	02/07/2021
					CT145	Cardiovascular devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	02/07/2021	02/07/2021
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	02/07/2021	02/07/2021
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	02/07/2021	02/07/2021
					CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	02/07/2021	02/07/2021
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	02/07/2021	02/07/2021
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	02/07/2021	02/07/2021
					CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	02/07/2021	02/07/2021
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	02/07/2021	02/07/2021
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	02/07/2021	02/07/2021
					CT233	Surgical	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	02/07/2021	02/07/2021
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	02/07/2021	02/07/2021
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	02/07/2021	02/07/2021
					CT2406	Implantable	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	02/07/2021	02/07/2021
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	02/07/2021	02/07/2021
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	02/07/2021	02/07/2021
					CT2407	Active implantable	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	02/07/2021	02/07/2021
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	02/07/2021	02/07/2021
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	02/07/2021	02/07/2021
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	02/07/2021	02/07/2021
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	02/07/2021	02/07/2021
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	02/07/2021	02/07/2021
					CT2727	Active implant devices/ancillaries	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	02/07/2021	02/07/2021
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	02/07/2021	02/07/2021
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	02/07/2021	02/07/2021
					CT2750	Implantable cardiac electrical stimulators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	02/07/2021	02/07/2021
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	02/07/2021	02/07/2021
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	02/07/2021	02/07/2021
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	02/07/2021	02/07/2021
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	02/07/2021	02/07/2021
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	02/07/2021	02/07/2021
					CT2985	Active	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	02/07/2021	02/07/2021
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	02/07/2021	02/07/2021
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	02/07/2021	02/07/2021
					CT3001	Implantable metal	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	02/07/2021	02/07/2021
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	02/07/2021	02/07/2021
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	02/07/2021	02/07/2021
					CT3002	Implantable synthetic polymer	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	02/07/2021	02/07/2021
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	02/07/2021	02/07/2021
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	02/07/2021	02/07/2021
					CT3006	Implantable inorganic material	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	02/07/2021	02/07/2021
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	02/07/2021	02/07/2021
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	02/07/2021	02/07/2021
					CT314	Electrically powered	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	02/07/2021	02/07/2021
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	02/07/2021	02/07/2021
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	02/07/2021	02/07/2021
					CT321	Long-term surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	02/07/2021	02/07/2021
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	02/07/2021	02/07/2021
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	02/07/2021	02/07/2021
					CT449	Generators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	02/07/2021	02/07/2021
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	02/07/2021	02/07/2021
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	02/07/2021	02/07/2021
					CT981	Single use	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	02/07/2021	02/07/2021
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	02/07/2021	02/07/2021
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	02/07/2021	02/07/2021
					CT983	Surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	02/07/2021	02/07/2021
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	02/07/2021	02/07/2021
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	02/07/2021	02/07/2021
11000	2021	All other reportable incidents	17882	Physiologic-monitoring defibrillation system	CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A0719	Unexpected Shutdown		

							F27	Problem Identified During Non-Clinical Procedure	A0719	Unexpected Shutdown		
11001	2021	All other reportable incidents	35852	Single-chamber implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A070909	Oversensing		
									A070910	Undersensing		
									A23	Use of Device Problem		
							F26	No Health Consequences or Impact	A070909	Oversensing		
									A070910	Undersensing		
									A23	Use of Device Problem		
11002	2021	All other reportable incidents	47267	Single-chamber implantable pacemaker, rate-responsive	CT141	Electrodes/Leads and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem	20/01/2022	20/01/2022
									A070504	Premature Discharge of Battery	20/01/2022	20/01/2022
							F08	Hospitalization or Prolonged Hospitalization	A0705	Battery Problem	20/01/2022	20/01/2022
									A070504	Premature Discharge of Battery	20/01/2022	20/01/2022
							F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A0705	Battery Problem	20/01/2022	20/01/2022
									A070504	Premature Discharge of Battery	20/01/2022	20/01/2022
							F1905	Device Revision or Replacement	A0705	Battery Problem	20/01/2022	20/01/2022
									A070504	Premature Discharge of Battery	20/01/2022	20/01/2022
					CT145	Cardiovascular devices	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem	20/01/2022	20/01/2022
									A070504	Premature Discharge of Battery	20/01/2022	20/01/2022
							F08	Hospitalization or Prolonged Hospitalization	A0705	Battery Problem	20/01/2022	20/01/2022
									A070504	Premature Discharge of Battery	20/01/2022	20/01/2022
							F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A0705	Battery Problem	20/01/2022	20/01/2022
									A070504	Premature Discharge of Battery	20/01/2022	20/01/2022
							F1905	Device Revision or Replacement	A0705	Battery Problem	20/01/2022	20/01/2022
									A070504	Premature Discharge of Battery	20/01/2022	20/01/2022
					CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem	20/01/2022	20/01/2022
									A070504	Premature Discharge of Battery	20/01/2022	20/01/2022
							F08	Hospitalization or Prolonged Hospitalization	A0705	Battery Problem	20/01/2022	20/01/2022
									A070504	Premature Discharge of Battery	20/01/2022	20/01/2022
							F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A0705	Battery Problem	20/01/2022	20/01/2022
									A070504	Premature Discharge of Battery	20/01/2022	20/01/2022
							F1905	Device Revision or Replacement	A0705	Battery Problem	20/01/2022	20/01/2022
									A070504	Premature Discharge of Battery	20/01/2022	20/01/2022
					CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem	20/01/2022	20/01/2022
									A070504	Premature Discharge of Battery	20/01/2022	20/01/2022
							F08	Hospitalization or Prolonged Hospitalization	A0705	Battery Problem	20/01/2022	20/01/2022
									A070504	Premature Discharge of Battery	20/01/2022	20/01/2022
							F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A0705	Battery Problem	20/01/2022	20/01/2022
									A070504	Premature Discharge of Battery	20/01/2022	20/01/2022
							F1905	Device Revision or Replacement	A0705	Battery Problem	20/01/2022	20/01/2022
									A070504	Premature Discharge of Battery	20/01/2022	20/01/2022
					CT233	Surgical	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem	20/01/2022	20/01/2022
									A070504	Premature Discharge of Battery	20/01/2022	20/01/2022
							F08	Hospitalization or Prolonged Hospitalization	A0705	Battery Problem	20/01/2022	20/01/2022
									A070504	Premature Discharge of Battery	20/01/2022	20/01/2022
							F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A0705	Battery Problem	20/01/2022	20/01/2022
									A070504	Premature Discharge of Battery	20/01/2022	20/01/2022
							F1905	Device Revision or Replacement	A0705	Battery Problem	20/01/2022	20/01/2022
									A070504	Premature Discharge of Battery	20/01/2022	20/01/2022
					CT2406	Implantable	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem	20/01/2022	20/01/2022
									A070504	Premature Discharge of Battery	20/01/2022	20/01/2022
							F08	Hospitalization or Prolonged Hospitalization	A0705	Battery Problem	20/01/2022	20/01/2022
									A070504	Premature Discharge of Battery	20/01/2022	20/01/2022
							F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A0705	Battery Problem	20/01/2022	20/01/2022
									A070504	Premature Discharge of Battery	20/01/2022	20/01/2022
							F1905	Device Revision or Replacement	A0705	Battery Problem	20/01/2022	20/01/2022
									A070504	Premature Discharge of Battery	20/01/2022	20/01/2022
					CT2407	Active implantable	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem	20/01/2022	20/01/2022
									A070504	Premature Discharge of Battery	20/01/2022	20/01/2022
							F08	Hospitalization or Prolonged Hospitalization	A0705	Battery Problem	20/01/2022	20/01/2022
									A070504	Premature Discharge of Battery	20/01/2022	20/01/2022
							F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A0705	Battery Problem	20/01/2022	20/01/2022
									A070504	Premature Discharge of Battery	20/01/2022	20/01/2022
							F1905	Device Revision or Replacement	A0705	Battery Problem	20/01/2022	20/01/2022
									A070504	Premature Discharge of Battery	20/01/2022	20/01/2022
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem	20/01/2022	20/01/2022
									A070504	Premature Discharge of Battery	20/01/2022	20/01/2022
							F08	Hospitalization or Prolonged Hospitalization	A0705	Battery Problem	20/01/2022	20/01/2022
									A070504	Premature Discharge of Battery	20/01/2022	20/01/2022
							F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A0705	Battery Problem	20/01/2022	20/01/2022
									A070504	Premature Discharge of Battery	20/01/2022	20/01/2022
							F1905	Device Revision or Replacement	A0705	Battery Problem	20/01/2022	20/01/2022
									A070504	Premature Discharge of Battery	20/01/2022	20/01/2022
					CT2727	Active implant devices/ancillaries	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem	20/01/2022	20/01/2022
									A070504	Premature Discharge of Battery	20/01/2022	20/01/2022
							F08	Hospitalization or Prolonged Hospitalization	A0705	Battery Problem	20/01/2022	20/01/2022
									A070504	Premature Discharge of Battery	20/01/2022	20/01/2022
							F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A0705	Battery Problem	20/01/2022	20/01/2022
									A070504	Premature Discharge of Battery	20/01/2022	20/01/2022
							F1905	Device Revision or Replacement	A0705	Battery Problem	20/01/2022	20/01/2022
									A070504	Premature Discharge of Battery	20/01/2022	20/01/2022
					CT2750	Implantable cardiac electrical stimulators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem	20/01/2022	20/01/2022
									A070504	Premature Discharge of Battery	20/01/2022	20/01/2022
							F08	Hospitalization or Prolonged Hospitalization	A0705	Battery Problem	20/01/2022	20/01/2022
									A070504	Premature Discharge of Battery	20/01/2022	20/01/2022
							F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A0705	Battery Problem	20/01/2022	20/01/2022
									A070504	Premature Discharge of Battery	20/01/2022	20/01/2022
							F1905	Device Revision or Replacement	A0705	Battery Problem	20/01/2022	20/01/2022
									A070504	Premature Discharge of Battery	20/01/2022	20/01/2022
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem	20/01/2022	20/01/2022

									A070504	Premature Discharge of Battery	20/01/2022	20/01/2022
							F08	Hospitalization or Prolonged Hospitalization	A0705	Battery Problem	20/01/2022	20/01/2022
									A070504	Premature Discharge of Battery	20/01/2022	20/01/2022
							F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A0705	Battery Problem	20/01/2022	20/01/2022
									A070504	Premature Discharge of Battery	20/01/2022	20/01/2022
							F1905	Device Revision or Replacement	A0705	Battery Problem	20/01/2022	20/01/2022
									A070504	Premature Discharge of Battery	20/01/2022	20/01/2022
					CT2985	Active	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem	20/01/2022	20/01/2022
									A070504	Premature Discharge of Battery	20/01/2022	20/01/2022
							F08	Hospitalization or Prolonged Hospitalization	A0705	Battery Problem	20/01/2022	20/01/2022
									A070504	Premature Discharge of Battery	20/01/2022	20/01/2022
							F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A0705	Battery Problem	20/01/2022	20/01/2022
									A070504	Premature Discharge of Battery	20/01/2022	20/01/2022
							F1905	Device Revision or Replacement	A0705	Battery Problem	20/01/2022	20/01/2022
									A070504	Premature Discharge of Battery	20/01/2022	20/01/2022
					CT3001	Implantable metal	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem	20/01/2022	20/01/2022
									A070504	Premature Discharge of Battery	20/01/2022	20/01/2022
							F08	Hospitalization or Prolonged Hospitalization	A0705	Battery Problem	20/01/2022	20/01/2022
									A070504	Premature Discharge of Battery	20/01/2022	20/01/2022
							F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A0705	Battery Problem	20/01/2022	20/01/2022
									A070504	Premature Discharge of Battery	20/01/2022	20/01/2022
							F1905	Device Revision or Replacement	A0705	Battery Problem	20/01/2022	20/01/2022
									A070504	Premature Discharge of Battery	20/01/2022	20/01/2022
					CT3002	Implantable synthetic polymer	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem	20/01/2022	20/01/2022
									A070504	Premature Discharge of Battery	20/01/2022	20/01/2022
							F08	Hospitalization or Prolonged Hospitalization	A0705	Battery Problem	20/01/2022	20/01/2022
									A070504	Premature Discharge of Battery	20/01/2022	20/01/2022
							F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A0705	Battery Problem	20/01/2022	20/01/2022
									A070504	Premature Discharge of Battery	20/01/2022	20/01/2022
							F1905	Device Revision or Replacement	A0705	Battery Problem	20/01/2022	20/01/2022
									A070504	Premature Discharge of Battery	20/01/2022	20/01/2022
					CT3006	Implantable inorganic material	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem	20/01/2022	20/01/2022
									A070504	Premature Discharge of Battery	20/01/2022	20/01/2022
							F08	Hospitalization or Prolonged Hospitalization	A0705	Battery Problem	20/01/2022	20/01/2022
									A070504	Premature Discharge of Battery	20/01/2022	20/01/2022
							F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A0705	Battery Problem	20/01/2022	20/01/2022
									A070504	Premature Discharge of Battery	20/01/2022	20/01/2022
							F1905	Device Revision or Replacement	A0705	Battery Problem	20/01/2022	20/01/2022
									A070504	Premature Discharge of Battery	20/01/2022	20/01/2022
					CT314	Electrically powered	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem	20/01/2022	20/01/2022
									A070504	Premature Discharge of Battery	20/01/2022	20/01/2022
							F08	Hospitalization or Prolonged Hospitalization	A0705	Battery Problem	20/01/2022	20/01/2022
									A070504	Premature Discharge of Battery	20/01/2022	20/01/2022
							F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A0705	Battery Problem	20/01/2022	20/01/2022
									A070504	Premature Discharge of Battery	20/01/2022	20/01/2022
							F1905	Device Revision or Replacement	A0705	Battery Problem	20/01/2022	20/01/2022
									A070504	Premature Discharge of Battery	20/01/2022	20/01/2022
					CT321	Long-term surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem	20/01/2022	20/01/2022
									A070504	Premature Discharge of Battery	20/01/2022	20/01/2022
							F08	Hospitalization or Prolonged Hospitalization	A0705	Battery Problem	20/01/2022	20/01/2022
									A070504	Premature Discharge of Battery	20/01/2022	20/01/2022
							F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A0705	Battery Problem	20/01/2022	20/01/2022
									A070504	Premature Discharge of Battery	20/01/2022	20/01/2022
							F1905	Device Revision or Replacement	A0705	Battery Problem	20/01/2022	20/01/2022
									A070504	Premature Discharge of Battery	20/01/2022	20/01/2022
					CT449	Generators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem	20/01/2022	20/01/2022
									A070504	Premature Discharge of Battery	20/01/2022	20/01/2022
							F08	Hospitalization or Prolonged Hospitalization	A0705	Battery Problem	20/01/2022	20/01/2022
									A070504	Premature Discharge of Battery	20/01/2022	20/01/2022
							F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A0705	Battery Problem	20/01/2022	20/01/2022
									A070504	Premature Discharge of Battery	20/01/2022	20/01/2022
							F1905	Device Revision or Replacement	A0705	Battery Problem	20/01/2022	20/01/2022
									A070504	Premature Discharge of Battery	20/01/2022	20/01/2022
					CT981	Single use	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem	20/01/2022	20/01/2022
									A070504	Premature Discharge of Battery	20/01/2022	20/01/2022
							F08	Hospitalization or Prolonged Hospitalization	A0705	Battery Problem	20/01/2022	20/01/2022
									A070504	Premature Discharge of Battery	20/01/2022	20/01/2022
							F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A0705	Battery Problem	20/01/2022	20/01/2022
									A070504	Premature Discharge of Battery	20/01/2022	20/01/2022
							F1905	Device Revision or Replacement	A0705	Battery Problem	20/01/2022	20/01/2022
									A070504	Premature Discharge of Battery	20/01/2022	20/01/2022
					CT983	Surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem	20/01/2022	20/01/2022
									A070504	Premature Discharge of Battery	20/01/2022	20/01/2022
							F08	Hospitalization or Prolonged Hospitalization	A0705	Battery Problem	20/01/2022	20/01/2022
									A070504	Premature Discharge of Battery	20/01/2022	20/01/2022
							F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A0705	Battery Problem	20/01/2022	20/01/2022
									A070504	Premature Discharge of Battery	20/01/2022	20/01/2022
							F1905	Device Revision or Replacement	A0705	Battery Problem	20/01/2022	20/01/2022
									A070504	Premature Discharge of Battery	20/01/2022	20/01/2022
11003	2021	All other reportable incidents	36241	Pacemaker/defibrillator lead	CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070909	Oversensing	26/04/2021	26/04/2021
									A071302	Inappropriate/Inadequate Shock/Stimulation	26/04/2021	26/04/2021
							F08	Hospitalization or Prolonged Hospitalization	A070909	Oversensing	26/04/2021	26/04/2021
									A071302	Inappropriate/Inadequate Shock/Stimulation	26/04/2021	26/04/2021
							F1905	Device Revision or Replacement	A070909	Oversensing	26/04/2021	26/04/2021
									A071302	Inappropriate/Inadequate Shock/Stimulation	26/04/2021	26/04/2021
11004	2021	All other reportable incidents	35852	Single-chamber implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	26/07/2021	26/07/2021
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	26/07/2021	26/07/2021
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	26/07/2021	26/07/2021

11005	2021	All other reportable incidents	35852	Single-chamber implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E2104	Electric Shock	A070909	Oversensing		
									A071302	Inappropriate/Inadequate Shock/Stimulation		
							F26	No Health Consequences or Impact	A070909	Oversensing		
									A071302	Inappropriate/Inadequate Shock/Stimulation		
11006	2021	All other reportable incidents	37265	Dual-chamber implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	21/04/2021	21/04/2021
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	21/04/2021	21/04/2021
							F12	Serious Injury/ Illness/ Impairment	A070504	Premature Discharge of Battery	21/04/2021	21/04/2021
							F19	Surgical Intervention	A070504	Premature Discharge of Battery	21/04/2021	21/04/2021
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	21/04/2021	21/04/2021
11007	2021	All other reportable incidents	47265	Dual-chamber implantable pacemaker, rate-responsive	CT141	Electrodes/Leads and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	04/06/2021	04/06/2021
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	04/06/2021	04/06/2021
							F1903	Device Explantation	A070504	Premature Discharge of Battery	04/06/2021	04/06/2021
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	04/06/2021	04/06/2021
					CT145	Cardiovascular devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	04/06/2021	04/06/2021
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	04/06/2021	04/06/2021
							F1903	Device Explantation	A070504	Premature Discharge of Battery	04/06/2021	04/06/2021
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	04/06/2021	04/06/2021
					CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	04/06/2021	04/06/2021
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	04/06/2021	04/06/2021
							F1903	Device Explantation	A070504	Premature Discharge of Battery	04/06/2021	04/06/2021
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	04/06/2021	04/06/2021
					CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	04/06/2021	04/06/2021
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	04/06/2021	04/06/2021
							F1903	Device Explantation	A070504	Premature Discharge of Battery	04/06/2021	04/06/2021
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	04/06/2021	04/06/2021
					CT233	Surgical	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	04/06/2021	04/06/2021
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	04/06/2021	04/06/2021
							F1903	Device Explantation	A070504	Premature Discharge of Battery	04/06/2021	04/06/2021
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	04/06/2021	04/06/2021
					CT2406	Implantable	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	04/06/2021	04/06/2021
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	04/06/2021	04/06/2021
							F1903	Device Explantation	A070504	Premature Discharge of Battery	04/06/2021	04/06/2021
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	04/06/2021	04/06/2021
					CT2407	Active implantable	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	04/06/2021	04/06/2021
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	04/06/2021	04/06/2021
							F1903	Device Explantation	A070504	Premature Discharge of Battery	04/06/2021	04/06/2021
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	04/06/2021	04/06/2021
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	04/06/2021	04/06/2021
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	04/06/2021	04/06/2021
							F1903	Device Explantation	A070504	Premature Discharge of Battery	04/06/2021	04/06/2021
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	04/06/2021	04/06/2021
					CT2727	Active implant devices/ancillaries	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	04/06/2021	04/06/2021
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	04/06/2021	04/06/2021
							F1903	Device Explantation	A070504	Premature Discharge of Battery	04/06/2021	04/06/2021
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	04/06/2021	04/06/2021
					CT2750	Implantable cardiac electrical stimulators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	04/06/2021	04/06/2021
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	04/06/2021	04/06/2021
							F1903	Device Explantation	A070504	Premature Discharge of Battery	04/06/2021	04/06/2021
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	04/06/2021	04/06/2021
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	04/06/2021	04/06/2021
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	04/06/2021	04/06/2021
							F1903	Device Explantation	A070504	Premature Discharge of Battery	04/06/2021	04/06/2021
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	04/06/2021	04/06/2021
					CT2985	Active	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	04/06/2021	04/06/2021
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	04/06/2021	04/06/2021
							F1903	Device Explantation	A070504	Premature Discharge of Battery	04/06/2021	04/06/2021
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	04/06/2021	04/06/2021
					CT3001	Implantable metal	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	04/06/2021	04/06/2021
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	04/06/2021	04/06/2021
							F1903	Device Explantation	A070504	Premature Discharge of Battery	04/06/2021	04/06/2021
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	04/06/2021	04/06/2021
					CT3002	Implantable synthetic polymer	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	04/06/2021	04/06/2021
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	04/06/2021	04/06/2021
							F1903	Device Explantation	A070504	Premature Discharge of Battery	04/06/2021	04/06/2021
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	04/06/2021	04/06/2021
					CT3006	Implantable inorganic material	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	04/06/2021	04/06/2021
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	04/06/2021	04/06/2021
							F1903	Device Explantation	A070504	Premature Discharge of Battery	04/06/2021	04/06/2021
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	04/06/2021	04/06/2021
					CT314	Electrically powered	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	04/06/2021	04/06/2021
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	04/06/2021	04/06/2021
							F1903	Device Explantation	A070504	Premature Discharge of Battery	04/06/2021	04/06/2021
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	04/06/2021	04/06/2021
					CT321	Long-term surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	04/06/2021	04/06/2021
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	04/06/2021	04/06/2021
							F1903	Device Explantation	A070504	Premature Discharge of Battery	04/06/2021	04/06/2021
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	04/06/2021	04/06/2021
					CT449	Generators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	04/06/2021	04/06/2021
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	04/06/2021	04/06/2021
							F1903	Device Explantation	A070504	Premature Discharge of Battery	04/06/2021	04/06/2021
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	04/06/2021	04/06/2021
					CT981	Single use	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	04/06/2021	04/06/2021
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	04/06/2021	04/06/2021
							F1903	Device Explantation	A070504	Premature Discharge of Battery	04/06/2021	04/06/2021
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	04/06/2021	04/06/2021
					CT983	Surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	04/06/2021	04/06/2021
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	04/06/2021	04/06/2021
							F1903	Device Explantation	A070504	Premature Discharge of Battery	04/06/2021	04/06/2021

							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	04/06/2021	04/06/2021
11008	2021	All other reportable incidents	36241	Pacemaker/defibrillator lead	CT1684	Cardiac pacemakers and associated devices	E2401	Insufficient Information	A072201	High Impedance		
							E2403	No Clinical Signs, Symptoms or Conditions	A072201	High Impedance		
							F26	No Health Consequences or Impact	A072201	High Impedance		
11009	2021	All other reportable incidents	37265	Dual-chamber implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E2401	Insufficient Information	A072201	High Impedance		
							E2403	No Clinical Signs, Symptoms or Conditions	A072201	High Impedance		
							F26	No Health Consequences or Impact	A072201	High Impedance		
11010	2021	All other reportable incidents	47265	Dual-chamber implantable pacemaker, rate-responsive	CT141	Electrodes/Leads and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem	29/04/2021	29/04/2021
									A070504	Premature Discharge of Battery	29/04/2021	29/04/2021
							F08	Hospitalization or Prolonged Hospitalization	A0705	Battery Problem	29/04/2021	29/04/2021
									A070504	Premature Discharge of Battery	29/04/2021	29/04/2021
							F1905	Device Revision or Replacement	A0705	Battery Problem	29/04/2021	29/04/2021
									A070504	Premature Discharge of Battery	29/04/2021	29/04/2021
					CT145	Cardiovascular devices	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem	29/04/2021	29/04/2021
									A070504	Premature Discharge of Battery	29/04/2021	29/04/2021
							F08	Hospitalization or Prolonged Hospitalization	A0705	Battery Problem	29/04/2021	29/04/2021
									A070504	Premature Discharge of Battery	29/04/2021	29/04/2021
							F1905	Device Revision or Replacement	A0705	Battery Problem	29/04/2021	29/04/2021
									A070504	Premature Discharge of Battery	29/04/2021	29/04/2021
					CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem	29/04/2021	29/04/2021
									A070504	Premature Discharge of Battery	29/04/2021	29/04/2021
							F08	Hospitalization or Prolonged Hospitalization	A0705	Battery Problem	29/04/2021	29/04/2021
									A070504	Premature Discharge of Battery	29/04/2021	29/04/2021
							F1905	Device Revision or Replacement	A0705	Battery Problem	29/04/2021	29/04/2021
									A070504	Premature Discharge of Battery	29/04/2021	29/04/2021
					CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem	29/04/2021	29/04/2021
									A070504	Premature Discharge of Battery	29/04/2021	29/04/2021
							F08	Hospitalization or Prolonged Hospitalization	A0705	Battery Problem	29/04/2021	29/04/2021
									A070504	Premature Discharge of Battery	29/04/2021	29/04/2021
							F1905	Device Revision or Replacement	A0705	Battery Problem	29/04/2021	29/04/2021
									A070504	Premature Discharge of Battery	29/04/2021	29/04/2021
					CT233	Surgical	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem	29/04/2021	29/04/2021
									A070504	Premature Discharge of Battery	29/04/2021	29/04/2021
							F08	Hospitalization or Prolonged Hospitalization	A0705	Battery Problem	29/04/2021	29/04/2021
									A070504	Premature Discharge of Battery	29/04/2021	29/04/2021
							F1905	Device Revision or Replacement	A0705	Battery Problem	29/04/2021	29/04/2021
									A070504	Premature Discharge of Battery	29/04/2021	29/04/2021
					CT2406	Implantable	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem	29/04/2021	29/04/2021
									A070504	Premature Discharge of Battery	29/04/2021	29/04/2021
							F08	Hospitalization or Prolonged Hospitalization	A0705	Battery Problem	29/04/2021	29/04/2021
									A070504	Premature Discharge of Battery	29/04/2021	29/04/2021
							F1905	Device Revision or Replacement	A0705	Battery Problem	29/04/2021	29/04/2021
									A070504	Premature Discharge of Battery	29/04/2021	29/04/2021
					CT2407	Active implantable	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem	29/04/2021	29/04/2021
									A070504	Premature Discharge of Battery	29/04/2021	29/04/2021
							F08	Hospitalization or Prolonged Hospitalization	A0705	Battery Problem	29/04/2021	29/04/2021
									A070504	Premature Discharge of Battery	29/04/2021	29/04/2021
							F1905	Device Revision or Replacement	A0705	Battery Problem	29/04/2021	29/04/2021
									A070504	Premature Discharge of Battery	29/04/2021	29/04/2021
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem	29/04/2021	29/04/2021
									A070504	Premature Discharge of Battery	29/04/2021	29/04/2021
							F08	Hospitalization or Prolonged Hospitalization	A0705	Battery Problem	29/04/2021	29/04/2021
									A070504	Premature Discharge of Battery	29/04/2021	29/04/2021
							F1905	Device Revision or Replacement	A0705	Battery Problem	29/04/2021	29/04/2021
									A070504	Premature Discharge of Battery	29/04/2021	29/04/2021
					CT2727	Active implant devices/ancillaries	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem	29/04/2021	29/04/2021
									A070504	Premature Discharge of Battery	29/04/2021	29/04/2021
							F08	Hospitalization or Prolonged Hospitalization	A0705	Battery Problem	29/04/2021	29/04/2021
									A070504	Premature Discharge of Battery	29/04/2021	29/04/2021
							F1905	Device Revision or Replacement	A0705	Battery Problem	29/04/2021	29/04/2021
									A070504	Premature Discharge of Battery	29/04/2021	29/04/2021
					CT2750	Implantable cardiac electrical stimulators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem	29/04/2021	29/04/2021
									A070504	Premature Discharge of Battery	29/04/2021	29/04/2021
							F08	Hospitalization or Prolonged Hospitalization	A0705	Battery Problem	29/04/2021	29/04/2021
									A070504	Premature Discharge of Battery	29/04/2021	29/04/2021
							F1905	Device Revision or Replacement	A0705	Battery Problem	29/04/2021	29/04/2021
									A070504	Premature Discharge of Battery	29/04/2021	29/04/2021
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem	29/04/2021	29/04/2021
									A070504	Premature Discharge of Battery	29/04/2021	29/04/2021
							F08	Hospitalization or Prolonged Hospitalization	A0705	Battery Problem	29/04/2021	29/04/2021
									A070504	Premature Discharge of Battery	29/04/2021	29/04/2021
							F1905	Device Revision or Replacement	A0705	Battery Problem	29/04/2021	29/04/2021
									A070504	Premature Discharge of Battery	29/04/2021	29/04/2021
					CT2985	Active	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem	29/04/2021	29/04/2021
									A070504	Premature Discharge of Battery	29/04/2021	29/04/2021
							F08	Hospitalization or Prolonged Hospitalization	A0705	Battery Problem	29/04/2021	29/04/2021
									A070504	Premature Discharge of Battery	29/04/2021	29/04/2021
							F1905	Device Revision or Replacement	A0705	Battery Problem	29/04/2021	29/04/2021
									A070504	Premature Discharge of Battery	29/04/2021	29/04/2021
					CT3001	Implantable metal	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem	29/04/2021	29/04/2021
									A070504	Premature Discharge of Battery	29/04/2021	29/04/2021
							F08	Hospitalization or Prolonged Hospitalization	A0705	Battery Problem	29/04/2021	29/04/2021
									A070504	Premature Discharge of Battery	29/04/2021	29/04/2021
							F1905	Device Revision or Replacement	A0705	Battery Problem	29/04/2021	29/04/2021
									A070504	Premature Discharge of Battery	29/04/2021	29/04/2021
					CT3002	Implantable synthetic polymer	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem	29/04/2021	29/04/2021
									A070504	Premature Discharge of Battery	29/04/2021	29/04/2021
							F08	Hospitalization or Prolonged Hospitalization	A0705	Battery Problem	29/04/2021	29/04/2021

									A070504	Premature Discharge of Battery	29/04/2021	29/04/2021
						F1905	Device Revision or Replacement	A0705	Battery Problem	29/04/2021	29/04/2021
								A070504	Premature Discharge of Battery	29/04/2021	29/04/2021
					CT3006	Implantable inorganic material	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem	29/04/2021	29/04/2021
									A070504	Premature Discharge of Battery	29/04/2021	29/04/2021
							F08	Hospitalization or Prolonged Hospitalization	A0705	Battery Problem	29/04/2021	29/04/2021
									A070504	Premature Discharge of Battery	29/04/2021	29/04/2021
							F1905	Device Revision or Replacement	A0705	Battery Problem	29/04/2021	29/04/2021
									A070504	Premature Discharge of Battery	29/04/2021	29/04/2021
					CT314	Electrically powered	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem	29/04/2021	29/04/2021
									A070504	Premature Discharge of Battery	29/04/2021	29/04/2021
							F08	Hospitalization or Prolonged Hospitalization	A0705	Battery Problem	29/04/2021	29/04/2021
									A070504	Premature Discharge of Battery	29/04/2021	29/04/2021
							F1905	Device Revision or Replacement	A0705	Battery Problem	29/04/2021	29/04/2021
									A070504	Premature Discharge of Battery	29/04/2021	29/04/2021
					CT321	Long-term surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem	29/04/2021	29/04/2021
									A070504	Premature Discharge of Battery	29/04/2021	29/04/2021
							F08	Hospitalization or Prolonged Hospitalization	A0705	Battery Problem	29/04/2021	29/04/2021
									A070504	Premature Discharge of Battery	29/04/2021	29/04/2021
							F1905	Device Revision or Replacement	A0705	Battery Problem	29/04/2021	29/04/2021
									A070504	Premature Discharge of Battery	29/04/2021	29/04/2021
					CT449	Generators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem	29/04/2021	29/04/2021
									A070504	Premature Discharge of Battery	29/04/2021	29/04/2021
							F08	Hospitalization or Prolonged Hospitalization	A0705	Battery Problem	29/04/2021	29/04/2021
									A070504	Premature Discharge of Battery	29/04/2021	29/04/2021
							F1905	Device Revision or Replacement	A0705	Battery Problem	29/04/2021	29/04/2021
									A070504	Premature Discharge of Battery	29/04/2021	29/04/2021
					CT981	Single use	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem	29/04/2021	29/04/2021
									A070504	Premature Discharge of Battery	29/04/2021	29/04/2021
							F08	Hospitalization or Prolonged Hospitalization	A0705	Battery Problem	29/04/2021	29/04/2021
									A070504	Premature Discharge of Battery	29/04/2021	29/04/2021
							F1905	Device Revision or Replacement	A0705	Battery Problem	29/04/2021	29/04/2021
									A070504	Premature Discharge of Battery	29/04/2021	29/04/2021
					CT983	Surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem	29/04/2021	29/04/2021
									A070504	Premature Discharge of Battery	29/04/2021	29/04/2021
							F08	Hospitalization or Prolonged Hospitalization	A0705	Battery Problem	29/04/2021	29/04/2021
									A070504	Premature Discharge of Battery	29/04/2021	29/04/2021
							F1905	Device Revision or Replacement	A0705	Battery Problem	29/04/2021	29/04/2021
									A070504	Premature Discharge of Battery	29/04/2021	29/04/2021
11011	2021	All other reportable incidents	47267	Single-chamber implantable pacemaker, rate-responsive	CT141	Electrodes/Leads and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem	21/04/2021	21/04/2021
									A070504	Premature Discharge of Battery	21/04/2021	21/04/2021
							F08	Hospitalization or Prolonged Hospitalization	A0705	Battery Problem	21/04/2021	21/04/2021
									A070504	Premature Discharge of Battery	21/04/2021	21/04/2021
							F12	Serious Injury/ Illness/ Impairment	A0705	Battery Problem	21/04/2021	21/04/2021
									A070504	Premature Discharge of Battery	21/04/2021	21/04/2021
							F1905	Device Revision or Replacement	A0705	Battery Problem	21/04/2021	21/04/2021
									A070504	Premature Discharge of Battery	21/04/2021	21/04/2021
					CT145	Cardiovascular devices	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem	21/04/2021	21/04/2021
									A070504	Premature Discharge of Battery	21/04/2021	21/04/2021
							F08	Hospitalization or Prolonged Hospitalization	A0705	Battery Problem	21/04/2021	21/04/2021
									A070504	Premature Discharge of Battery	21/04/2021	21/04/2021
							F12	Serious Injury/ Illness/ Impairment	A0705	Battery Problem	21/04/2021	21/04/2021
									A070504	Premature Discharge of Battery	21/04/2021	21/04/2021
							F1905	Device Revision or Replacement	A0705	Battery Problem	21/04/2021	21/04/2021
									A070504	Premature Discharge of Battery	21/04/2021	21/04/2021
					CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem	21/04/2021	21/04/2021
									A070504	Premature Discharge of Battery	21/04/2021	21/04/2021
							F08	Hospitalization or Prolonged Hospitalization	A0705	Battery Problem	21/04/2021	21/04/2021
									A070504	Premature Discharge of Battery	21/04/2021	21/04/2021
							F12	Serious Injury/ Illness/ Impairment	A0705	Battery Problem	21/04/2021	21/04/2021
									A070504	Premature Discharge of Battery	21/04/2021	21/04/2021
							F1905	Device Revision or Replacement	A0705	Battery Problem	21/04/2021	21/04/2021
									A070504	Premature Discharge of Battery	21/04/2021	21/04/2021
					CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem	21/04/2021	21/04/2021
									A070504	Premature Discharge of Battery	21/04/2021	21/04/2021
							F08	Hospitalization or Prolonged Hospitalization	A0705	Battery Problem	21/04/2021	21/04/2021
									A070504	Premature Discharge of Battery	21/04/2021	21/04/2021
							F12	Serious Injury/ Illness/ Impairment	A0705	Battery Problem	21/04/2021	21/04/2021
									A070504	Premature Discharge of Battery	21/04/2021	21/04/2021
							F1905	Device Revision or Replacement	A0705	Battery Problem	21/04/2021	21/04/2021
									A070504	Premature Discharge of Battery	21/04/2021	21/04/2021
					CT233	Surgical	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem	21/04/2021	21/04/2021
									A070504	Premature Discharge of Battery	21/04/2021	21/04/2021
							F08	Hospitalization or Prolonged Hospitalization	A0705	Battery Problem	21/04/2021	21/04/2021
									A070504	Premature Discharge of Battery	21/04/2021	21/04/2021
							F12	Serious Injury/ Illness/ Impairment	A0705	Battery Problem	21/04/2021	21/04/2021
									A070504	Premature Discharge of Battery	21/04/2021	21/04/2021
							F1905	Device Revision or Replacement	A0705	Battery Problem	21/04/2021	21/04/2021
									A070504	Premature Discharge of Battery	21/04/2021	21/04/2021
					CT2406	Implantable	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem	21/04/2021	21/04/2021
									A070504	Premature Discharge of Battery	21/04/2021	21/04/2021
							F08	Hospitalization or Prolonged Hospitalization	A0705	Battery Problem	21/04/2021	21/04/2021
									A070504	Premature Discharge of Battery	21/04/2021	21/04/2021
							F12	Serious Injury/ Illness/ Impairment	A0705	Battery Problem	21/04/2021	21/04/2021
									A070504	Premature Discharge of Battery	21/04/2021	21/04/2021
							F1905	Device Revision or Replacement	A0705	Battery Problem	21/04/2021	21/04/2021
									A070504	Premature Discharge of Battery	21/04/2021	21/04/2021
					CT2407	Active implantable	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem	21/04/2021	21/04/2021

								A070504	Premature Discharge of Battery	21/04/2021	21/04/2021
						F08	Hospitalization or Prolonged Hospitalization	A0705	Battery Problem	21/04/2021	21/04/2021
								A070504	Premature Discharge of Battery	21/04/2021	21/04/2021
						F12	Serious Injury/ Illness/ Impairment	A0705	Battery Problem	21/04/2021	21/04/2021
								A070504	Premature Discharge of Battery	21/04/2021	21/04/2021
						F1905	Device Revision or Replacement	A0705	Battery Problem	21/04/2021	21/04/2021
								A070504	Premature Discharge of Battery	21/04/2021	21/04/2021
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem	21/04/2021	21/04/2021
									A070504	Premature Discharge of Battery	21/04/2021	21/04/2021
							F08	Hospitalization or Prolonged Hospitalization	A0705	Battery Problem	21/04/2021	21/04/2021
									A070504	Premature Discharge of Battery	21/04/2021	21/04/2021
							F12	Serious Injury/ Illness/ Impairment	A0705	Battery Problem	21/04/2021	21/04/2021
									A070504	Premature Discharge of Battery	21/04/2021	21/04/2021
							F1905	Device Revision or Replacement	A0705	Battery Problem	21/04/2021	21/04/2021
									A070504	Premature Discharge of Battery	21/04/2021	21/04/2021
					CT2727	Active implant devices/ancillaries	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem	21/04/2021	21/04/2021
									A070504	Premature Discharge of Battery	21/04/2021	21/04/2021
							F08	Hospitalization or Prolonged Hospitalization	A0705	Battery Problem	21/04/2021	21/04/2021
									A070504	Premature Discharge of Battery	21/04/2021	21/04/2021
							F12	Serious Injury/ Illness/ Impairment	A0705	Battery Problem	21/04/2021	21/04/2021
									A070504	Premature Discharge of Battery	21/04/2021	21/04/2021
							F1905	Device Revision or Replacement	A0705	Battery Problem	21/04/2021	21/04/2021
									A070504	Premature Discharge of Battery	21/04/2021	21/04/2021
					CT2750	Implantable cardiac electrical stimulators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem	21/04/2021	21/04/2021
									A070504	Premature Discharge of Battery	21/04/2021	21/04/2021
							F08	Hospitalization or Prolonged Hospitalization	A0705	Battery Problem	21/04/2021	21/04/2021
									A070504	Premature Discharge of Battery	21/04/2021	21/04/2021
							F12	Serious Injury/ Illness/ Impairment	A0705	Battery Problem	21/04/2021	21/04/2021
									A070504	Premature Discharge of Battery	21/04/2021	21/04/2021
							F1905	Device Revision or Replacement	A0705	Battery Problem	21/04/2021	21/04/2021
									A070504	Premature Discharge of Battery	21/04/2021	21/04/2021
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem	21/04/2021	21/04/2021
									A070504	Premature Discharge of Battery	21/04/2021	21/04/2021
							F08	Hospitalization or Prolonged Hospitalization	A0705	Battery Problem	21/04/2021	21/04/2021
									A070504	Premature Discharge of Battery	21/04/2021	21/04/2021
							F12	Serious Injury/ Illness/ Impairment	A0705	Battery Problem	21/04/2021	21/04/2021
									A070504	Premature Discharge of Battery	21/04/2021	21/04/2021
							F1905	Device Revision or Replacement	A0705	Battery Problem	21/04/2021	21/04/2021
									A070504	Premature Discharge of Battery	21/04/2021	21/04/2021
					CT2985	Active	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem	21/04/2021	21/04/2021
									A070504	Premature Discharge of Battery	21/04/2021	21/04/2021
							F08	Hospitalization or Prolonged Hospitalization	A0705	Battery Problem	21/04/2021	21/04/2021
									A070504	Premature Discharge of Battery	21/04/2021	21/04/2021
							F12	Serious Injury/ Illness/ Impairment	A0705	Battery Problem	21/04/2021	21/04/2021
									A070504	Premature Discharge of Battery	21/04/2021	21/04/2021
							F1905	Device Revision or Replacement	A0705	Battery Problem	21/04/2021	21/04/2021
									A070504	Premature Discharge of Battery	21/04/2021	21/04/2021
					CT3001	Implantable metal	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem	21/04/2021	21/04/2021
									A070504	Premature Discharge of Battery	21/04/2021	21/04/2021
							F08	Hospitalization or Prolonged Hospitalization	A0705	Battery Problem	21/04/2021	21/04/2021
									A070504	Premature Discharge of Battery	21/04/2021	21/04/2021
							F12	Serious Injury/ Illness/ Impairment	A0705	Battery Problem	21/04/2021	21/04/2021
									A070504	Premature Discharge of Battery	21/04/2021	21/04/2021
							F1905	Device Revision or Replacement	A0705	Battery Problem	21/04/2021	21/04/2021
									A070504	Premature Discharge of Battery	21/04/2021	21/04/2021
					CT3002	Implantable synthetic polymer	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem	21/04/2021	21/04/2021
									A070504	Premature Discharge of Battery	21/04/2021	21/04/2021
							F08	Hospitalization or Prolonged Hospitalization	A0705	Battery Problem	21/04/2021	21/04/2021
									A070504	Premature Discharge of Battery	21/04/2021	21/04/2021
							F12	Serious Injury/ Illness/ Impairment	A0705	Battery Problem	21/04/2021	21/04/2021
									A070504	Premature Discharge of Battery	21/04/2021	21/04/2021
							F1905	Device Revision or Replacement	A0705	Battery Problem	21/04/2021	21/04/2021
									A070504	Premature Discharge of Battery	21/04/2021	21/04/2021
					CT3006	Implantable inorganic material	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem	21/04/2021	21/04/2021
									A070504	Premature Discharge of Battery	21/04/2021	21/04/2021
							F08	Hospitalization or Prolonged Hospitalization	A0705	Battery Problem	21/04/2021	21/04/2021
									A070504	Premature Discharge of Battery	21/04/2021	21/04/2021
							F12	Serious Injury/ Illness/ Impairment	A0705	Battery Problem	21/04/2021	21/04/2021
									A070504	Premature Discharge of Battery	21/04/2021	21/04/2021
							F1905	Device Revision or Replacement	A0705	Battery Problem	21/04/2021	21/04/2021
									A070504	Premature Discharge of Battery	21/04/2021	21/04/2021
					CT314	Electrically powered	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem	21/04/2021	21/04/2021
									A070504	Premature Discharge of Battery	21/04/2021	21/04/2021
							F08	Hospitalization or Prolonged Hospitalization	A0705	Battery Problem	21/04/2021	21/04/2021
									A070504	Premature Discharge of Battery	21/04/2021	21/04/2021
							F12	Serious Injury/ Illness/ Impairment	A0705	Battery Problem	21/04/2021	21/04/2021
									A070504	Premature Discharge of Battery	21/04/2021	21/04/2021
							F1905	Device Revision or Replacement	A0705	Battery Problem	21/04/2021	21/04/2021
									A070504	Premature Discharge of Battery	21/04/2021	21/04/2021
					CT321	Long-term surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem	21/04/2021	21/04/2021
									A070504	Premature Discharge of Battery	21/04/2021	21/04/2021
							F08	Hospitalization or Prolonged Hospitalization	A0705	Battery Problem	21/04/2021	21/04/2021
									A070504	Premature Discharge of Battery	21/04/2021	21/04/2021
							F12	Serious Injury/ Illness/ Impairment	A0705	Battery Problem	21/04/2021	21/04/2021
									A070504	Premature Discharge of Battery	21/04/2021	21/04/2021
							F1905	Device Revision or Replacement	A0705	Battery Problem	21/04/2021	21/04/2021
									A070504	Premature Discharge of Battery	21/04/2021	21/04/2021
					CT449	Generators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem	21/04/2021	21/04/2021

								A070504	Premature Discharge of Battery	21/04/2021	21/04/2021
						F08	Hospitalization or Prolonged Hospitalization	A0705	Battery Problem	21/04/2021	21/04/2021
								A070504	Premature Discharge of Battery	21/04/2021	21/04/2021
						F12	Serious Injury/ Illness/ Impairment	A0705	Battery Problem	21/04/2021	21/04/2021
								A070504	Premature Discharge of Battery	21/04/2021	21/04/2021
						F1905	Device Revision or Replacement	A0705	Battery Problem	21/04/2021	21/04/2021
								A070504	Premature Discharge of Battery	21/04/2021	21/04/2021
				CT981	Single use	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem	21/04/2021	21/04/2021
								A070504	Premature Discharge of Battery	21/04/2021	21/04/2021
						F08	Hospitalization or Prolonged Hospitalization	A0705	Battery Problem	21/04/2021	21/04/2021
								A070504	Premature Discharge of Battery	21/04/2021	21/04/2021
						F12	Serious Injury/ Illness/ Impairment	A0705	Battery Problem	21/04/2021	21/04/2021
								A070504	Premature Discharge of Battery	21/04/2021	21/04/2021
						F1905	Device Revision or Replacement	A0705	Battery Problem	21/04/2021	21/04/2021
								A070504	Premature Discharge of Battery	21/04/2021	21/04/2021
				CT983	Surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem	21/04/2021	21/04/2021
								A070504	Premature Discharge of Battery	21/04/2021	21/04/2021
						F08	Hospitalization or Prolonged Hospitalization	A0705	Battery Problem	21/04/2021	21/04/2021
								A070504	Premature Discharge of Battery	21/04/2021	21/04/2021
						F12	Serious Injury/ Illness/ Impairment	A0705	Battery Problem	21/04/2021	21/04/2021
								A070504	Premature Discharge of Battery	21/04/2021	21/04/2021
						F1905	Device Revision or Replacement	A0705	Battery Problem	21/04/2021	21/04/2021
								A070504	Premature Discharge of Battery	21/04/2021	21/04/2021
11012	2021	All other reportable incidents	47265	Dual-chamber implantable pacemaker, rate-responsive	CT141	Electrodes/Leads and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	30/06/2021	30/06/2021
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	30/06/2021	30/06/2021
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	30/06/2021	30/06/2021
					CT145	Cardiovascular devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	30/06/2021	30/06/2021
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	30/06/2021	30/06/2021
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	30/06/2021	30/06/2021
					CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	30/06/2021	30/06/2021
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	30/06/2021	30/06/2021
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	30/06/2021	30/06/2021
					CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	30/06/2021	30/06/2021
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	30/06/2021	30/06/2021
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	30/06/2021	30/06/2021
					CT233	Surgical	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	30/06/2021	30/06/2021
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	30/06/2021	30/06/2021
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	30/06/2021	30/06/2021
					CT2406	Implantable	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	30/06/2021	30/06/2021
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	30/06/2021	30/06/2021
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	30/06/2021	30/06/2021
					CT2407	Active implantable	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	30/06/2021	30/06/2021
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	30/06/2021	30/06/2021
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	30/06/2021	30/06/2021
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	30/06/2021	30/06/2021
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	30/06/2021	30/06/2021
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	30/06/2021	30/06/2021
					CT2727	Active implant devices/ancillaries	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	30/06/2021	30/06/2021
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	30/06/2021	30/06/2021
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	30/06/2021	30/06/2021
					CT2750	Implantable cardiac electrical stimulators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	30/06/2021	30/06/2021
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	30/06/2021	30/06/2021
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	30/06/2021	30/06/2021
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	30/06/2021	30/06/2021
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	30/06/2021	30/06/2021
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	30/06/2021	30/06/2021
					CT2985	Active	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	30/06/2021	30/06/2021
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	30/06/2021	30/06/2021
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	30/06/2021	30/06/2021
					CT3001	Implantable metal	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	30/06/2021	30/06/2021
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	30/06/2021	30/06/2021
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	30/06/2021	30/06/2021
					CT3002	Implantable synthetic polymer	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	30/06/2021	30/06/2021
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	30/06/2021	30/06/2021
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	30/06/2021	30/06/2021
					CT3006	Implantable inorganic material	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	30/06/2021	30/06/2021
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	30/06/2021	30/06/2021
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	30/06/2021	30/06/2021
					CT314	Electrically powered	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	30/06/2021	30/06/2021
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	30/06/2021	30/06/2021
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	30/06/2021	30/06/2021
					CT321	Long-term surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	30/06/2021	30/06/2021
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	30/06/2021	30/06/2021
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	30/06/2021	30/06/2021
					CT449	Generators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	30/06/2021	30/06/2021
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	30/06/2021	30/06/2021
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	30/06/2021	30/06/2021
					CT981	Single use	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	30/06/2021	30/06/2021
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	30/06/2021	30/06/2021
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	30/06/2021	30/06/2021
					CT983	Surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	30/06/2021	30/06/2021
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	30/06/2021	30/06/2021
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	30/06/2021	30/06/2021
11013	2021	All other reportable incidents	47265	Dual-chamber implantable pacemaker, rate-responsive	CT141	Electrodes/Leads and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem	07/07/2021	07/07/2021
									A070504	Premature Discharge of Battery	07/07/2021	07/07/2021
							F08	Hospitalization or Prolonged Hospitalization	A0705	Battery Problem	07/07/2021	07/07/2021
									A070504	Premature Discharge of Battery	07/07/2021	07/07/2021
							F1905	Device Revision or Replacement	A0705	Battery Problem	07/07/2021	07/07/2021

								A070504	Premature Discharge of Battery	07/07/2021	07/07/2021
					CT145	Cardiovascular devices	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem	07/07/2021	07/07/2021
									A070504	Premature Discharge of Battery	07/07/2021	07/07/2021
							F08	Hospitalization or Prolonged Hospitalization	A0705	Battery Problem	07/07/2021	07/07/2021
									A070504	Premature Discharge of Battery	07/07/2021	07/07/2021
							F1905	Device Revision or Replacement	A0705	Battery Problem	07/07/2021	07/07/2021
									A070504	Premature Discharge of Battery	07/07/2021	07/07/2021
					CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem	07/07/2021	07/07/2021
									A070504	Premature Discharge of Battery	07/07/2021	07/07/2021
							F08	Hospitalization or Prolonged Hospitalization	A0705	Battery Problem	07/07/2021	07/07/2021
									A070504	Premature Discharge of Battery	07/07/2021	07/07/2021
							F1905	Device Revision or Replacement	A0705	Battery Problem	07/07/2021	07/07/2021
									A070504	Premature Discharge of Battery	07/07/2021	07/07/2021
					CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem	07/07/2021	07/07/2021
									A070504	Premature Discharge of Battery	07/07/2021	07/07/2021
							F08	Hospitalization or Prolonged Hospitalization	A0705	Battery Problem	07/07/2021	07/07/2021
									A070504	Premature Discharge of Battery	07/07/2021	07/07/2021
							F1905	Device Revision or Replacement	A0705	Battery Problem	07/07/2021	07/07/2021
									A070504	Premature Discharge of Battery	07/07/2021	07/07/2021
					CT233	Surgical	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem	07/07/2021	07/07/2021
									A070504	Premature Discharge of Battery	07/07/2021	07/07/2021
							F08	Hospitalization or Prolonged Hospitalization	A0705	Battery Problem	07/07/2021	07/07/2021
									A070504	Premature Discharge of Battery	07/07/2021	07/07/2021
							F1905	Device Revision or Replacement	A0705	Battery Problem	07/07/2021	07/07/2021
									A070504	Premature Discharge of Battery	07/07/2021	07/07/2021
					CT2406	Implantable	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem	07/07/2021	07/07/2021
									A070504	Premature Discharge of Battery	07/07/2021	07/07/2021
							F08	Hospitalization or Prolonged Hospitalization	A0705	Battery Problem	07/07/2021	07/07/2021
									A070504	Premature Discharge of Battery	07/07/2021	07/07/2021
							F1905	Device Revision or Replacement	A0705	Battery Problem	07/07/2021	07/07/2021
									A070504	Premature Discharge of Battery	07/07/2021	07/07/2021
					CT2407	Active implantable	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem	07/07/2021	07/07/2021
									A070504	Premature Discharge of Battery	07/07/2021	07/07/2021
							F08	Hospitalization or Prolonged Hospitalization	A0705	Battery Problem	07/07/2021	07/07/2021
									A070504	Premature Discharge of Battery	07/07/2021	07/07/2021
							F1905	Device Revision or Replacement	A0705	Battery Problem	07/07/2021	07/07/2021
									A070504	Premature Discharge of Battery	07/07/2021	07/07/2021
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem	07/07/2021	07/07/2021
									A070504	Premature Discharge of Battery	07/07/2021	07/07/2021
							F08	Hospitalization or Prolonged Hospitalization	A0705	Battery Problem	07/07/2021	07/07/2021
									A070504	Premature Discharge of Battery	07/07/2021	07/07/2021
							F1905	Device Revision or Replacement	A0705	Battery Problem	07/07/2021	07/07/2021
									A070504	Premature Discharge of Battery	07/07/2021	07/07/2021
					CT2727	Active implant devices/ancillaries	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem	07/07/2021	07/07/2021
									A070504	Premature Discharge of Battery	07/07/2021	07/07/2021
							F08	Hospitalization or Prolonged Hospitalization	A0705	Battery Problem	07/07/2021	07/07/2021
									A070504	Premature Discharge of Battery	07/07/2021	07/07/2021
							F1905	Device Revision or Replacement	A0705	Battery Problem	07/07/2021	07/07/2021
									A070504	Premature Discharge of Battery	07/07/2021	07/07/2021
					CT2750	Implantable cardiac electrical stimulators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem	07/07/2021	07/07/2021
									A070504	Premature Discharge of Battery	07/07/2021	07/07/2021
							F08	Hospitalization or Prolonged Hospitalization	A0705	Battery Problem	07/07/2021	07/07/2021
									A070504	Premature Discharge of Battery	07/07/2021	07/07/2021
							F1905	Device Revision or Replacement	A0705	Battery Problem	07/07/2021	07/07/2021
									A070504	Premature Discharge of Battery	07/07/2021	07/07/2021
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem	07/07/2021	07/07/2021
									A070504	Premature Discharge of Battery	07/07/2021	07/07/2021
							F08	Hospitalization or Prolonged Hospitalization	A0705	Battery Problem	07/07/2021	07/07/2021
									A070504	Premature Discharge of Battery	07/07/2021	07/07/2021
							F1905	Device Revision or Replacement	A0705	Battery Problem	07/07/2021	07/07/2021
									A070504	Premature Discharge of Battery	07/07/2021	07/07/2021
					CT2985	Active	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem	07/07/2021	07/07/2021
									A070504	Premature Discharge of Battery	07/07/2021	07/07/2021
							F08	Hospitalization or Prolonged Hospitalization	A0705	Battery Problem	07/07/2021	07/07/2021
									A070504	Premature Discharge of Battery	07/07/2021	07/07/2021
							F1905	Device Revision or Replacement	A0705	Battery Problem	07/07/2021	07/07/2021
									A070504	Premature Discharge of Battery	07/07/2021	07/07/2021
					CT3001	Implantable metal	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem	07/07/2021	07/07/2021
									A070504	Premature Discharge of Battery	07/07/2021	07/07/2021
							F08	Hospitalization or Prolonged Hospitalization	A0705	Battery Problem	07/07/2021	07/07/2021
									A070504	Premature Discharge of Battery	07/07/2021	07/07/2021
							F1905	Device Revision or Replacement	A0705	Battery Problem	07/07/2021	07/07/2021
									A070504	Premature Discharge of Battery	07/07/2021	07/07/2021
					CT3002	Implantable synthetic polymer	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem	07/07/2021	07/07/2021
									A070504	Premature Discharge of Battery	07/07/2021	07/07/2021
							F08	Hospitalization or Prolonged Hospitalization	A0705	Battery Problem	07/07/2021	07/07/2021
									A070504	Premature Discharge of Battery	07/07/2021	07/07/2021
							F1905	Device Revision or Replacement	A0705	Battery Problem	07/07/2021	07/07/2021
									A070504	Premature Discharge of Battery	07/07/2021	07/07/2021
					CT3006	Implantable inorganic material	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem	07/07/2021	07/07/2021
									A070504	Premature Discharge of Battery	07/07/2021	07/07/2021
							F08	Hospitalization or Prolonged Hospitalization	A0705	Battery Problem	07/07/2021	07/07/2021
									A070504	Premature Discharge of Battery	07/07/2021	07/07/2021
							F1905	Device Revision or Replacement	A0705	Battery Problem	07/07/2021	07/07/2021
									A070504	Premature Discharge of Battery	07/07/2021	07/07/2021
					CT314	Electrically powered	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem	07/07/2021	07/07/2021
									A070504	Premature Discharge of Battery	07/07/2021	07/07/2021
							F08	Hospitalization or Prolonged Hospitalization	A0705	Battery Problem	07/07/2021	07/07/2021

								A070504	Premature Discharge of Battery	07/07/2021	07/07/2021
						F1905	Device Revision or Replacement	A0705	Battery Problem	07/07/2021	07/07/2021
								A070504	Premature Discharge of Battery	07/07/2021	07/07/2021
				CT321	Long-term surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem	07/07/2021	07/07/2021
								A070504	Premature Discharge of Battery	07/07/2021	07/07/2021
						F08	Hospitalization or Prolonged Hospitalization	A0705	Battery Problem	07/07/2021	07/07/2021
								A070504	Premature Discharge of Battery	07/07/2021	07/07/2021
						F1905	Device Revision or Replacement	A0705	Battery Problem	07/07/2021	07/07/2021
								A070504	Premature Discharge of Battery	07/07/2021	07/07/2021
				CT449	Generators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem	07/07/2021	07/07/2021
								A070504	Premature Discharge of Battery	07/07/2021	07/07/2021
						F08	Hospitalization or Prolonged Hospitalization	A0705	Battery Problem	07/07/2021	07/07/2021
								A070504	Premature Discharge of Battery	07/07/2021	07/07/2021
						F1905	Device Revision or Replacement	A0705	Battery Problem	07/07/2021	07/07/2021
								A070504	Premature Discharge of Battery	07/07/2021	07/07/2021
				CT981	Single use	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem	07/07/2021	07/07/2021
								A070504	Premature Discharge of Battery	07/07/2021	07/07/2021
						F08	Hospitalization or Prolonged Hospitalization	A0705	Battery Problem	07/07/2021	07/07/2021
								A070504	Premature Discharge of Battery	07/07/2021	07/07/2021
						F1905	Device Revision or Replacement	A0705	Battery Problem	07/07/2021	07/07/2021
								A070504	Premature Discharge of Battery	07/07/2021	07/07/2021
				CT983	Surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem	07/07/2021	07/07/2021
								A070504	Premature Discharge of Battery	07/07/2021	07/07/2021
						F08	Hospitalization or Prolonged Hospitalization	A0705	Battery Problem	07/07/2021	07/07/2021
								A070504	Premature Discharge of Battery	07/07/2021	07/07/2021
						F1905	Device Revision or Replacement	A0705	Battery Problem	07/07/2021	07/07/2021
								A070504	Premature Discharge of Battery	07/07/2021	07/07/2021
11014	2021	All other reportable incidents	47265	Dual-chamber implantable pacemaker, rate-responsive	CT141	Electrodes/Leads and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem	26/05/2021	26/05/2021
									A070504	Premature Discharge of Battery	26/05/2021	26/05/2021
						F08	Hospitalization or Prolonged Hospitalization	A0705	Battery Problem	26/05/2021	26/05/2021
								A070504	Premature Discharge of Battery	26/05/2021	26/05/2021
						F1905	Device Revision or Replacement	A0705	Battery Problem	26/05/2021	26/05/2021
								A070504	Premature Discharge of Battery	26/05/2021	26/05/2021
				CT145	Cardiovascular devices	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem	26/05/2021	26/05/2021
								A070504	Premature Discharge of Battery	26/05/2021	26/05/2021
						F08	Hospitalization or Prolonged Hospitalization	A0705	Battery Problem	26/05/2021	26/05/2021
								A070504	Premature Discharge of Battery	26/05/2021	26/05/2021
						F1905	Device Revision or Replacement	A0705	Battery Problem	26/05/2021	26/05/2021
								A070504	Premature Discharge of Battery	26/05/2021	26/05/2021
				CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem	26/05/2021	26/05/2021
								A070504	Premature Discharge of Battery	26/05/2021	26/05/2021
						F08	Hospitalization or Prolonged Hospitalization	A0705	Battery Problem	26/05/2021	26/05/2021
								A070504	Premature Discharge of Battery	26/05/2021	26/05/2021
						F1905	Device Revision or Replacement	A0705	Battery Problem	26/05/2021	26/05/2021
								A070504	Premature Discharge of Battery	26/05/2021	26/05/2021
				CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem	26/05/2021	26/05/2021
								A070504	Premature Discharge of Battery	26/05/2021	26/05/2021
						F08	Hospitalization or Prolonged Hospitalization	A0705	Battery Problem	26/05/2021	26/05/2021
								A070504	Premature Discharge of Battery	26/05/2021	26/05/2021
						F1905	Device Revision or Replacement	A0705	Battery Problem	26/05/2021	26/05/2021
								A070504	Premature Discharge of Battery	26/05/2021	26/05/2021
				CT233	Surgical	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem	26/05/2021	26/05/2021
								A070504	Premature Discharge of Battery	26/05/2021	26/05/2021
						F08	Hospitalization or Prolonged Hospitalization	A0705	Battery Problem	26/05/2021	26/05/2021
								A070504	Premature Discharge of Battery	26/05/2021	26/05/2021
						F1905	Device Revision or Replacement	A0705	Battery Problem	26/05/2021	26/05/2021
								A070504	Premature Discharge of Battery	26/05/2021	26/05/2021
				CT2406	Implantable	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem	26/05/2021	26/05/2021
								A070504	Premature Discharge of Battery	26/05/2021	26/05/2021
						F08	Hospitalization or Prolonged Hospitalization	A0705	Battery Problem	26/05/2021	26/05/2021
								A070504	Premature Discharge of Battery	26/05/2021	26/05/2021
						F1905	Device Revision or Replacement	A0705	Battery Problem	26/05/2021	26/05/2021
								A070504	Premature Discharge of Battery	26/05/2021	26/05/2021
				CT2407	Active implantable	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem	26/05/2021	26/05/2021
								A070504	Premature Discharge of Battery	26/05/2021	26/05/2021
						F08	Hospitalization or Prolonged Hospitalization	A0705	Battery Problem	26/05/2021	26/05/2021
								A070504	Premature Discharge of Battery	26/05/2021	26/05/2021
						F1905	Device Revision or Replacement	A0705	Battery Problem	26/05/2021	26/05/2021
								A070504	Premature Discharge of Battery	26/05/2021	26/05/2021
				CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem	26/05/2021	26/05/2021
								A070504	Premature Discharge of Battery	26/05/2021	26/05/2021
						F08	Hospitalization or Prolonged Hospitalization	A0705	Battery Problem	26/05/2021	26/05/2021
								A070504	Premature Discharge of Battery	26/05/2021	26/05/2021
						F1905	Device Revision or Replacement	A0705	Battery Problem	26/05/2021	26/05/2021
								A070504	Premature Discharge of Battery	26/05/2021	26/05/2021
				CT2727	Active implant devices/ancillaries	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem	26/05/2021	26/05/2021
								A070504	Premature Discharge of Battery	26/05/2021	26/05/2021
						F08	Hospitalization or Prolonged Hospitalization	A0705	Battery Problem	26/05/2021	26/05/2021
								A070504	Premature Discharge of Battery	26/05/2021	26/05/2021
						F1905	Device Revision or Replacement	A0705	Battery Problem	26/05/2021	26/05/2021
								A070504	Premature Discharge of Battery	26/05/2021	26/05/2021
				CT2750	Implantable cardiac electrical stimulators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem	26/05/2021	26/05/2021
								A070504	Premature Discharge of Battery	26/05/2021	26/05/2021
						F08	Hospitalization or Prolonged Hospitalization	A0705	Battery Problem	26/05/2021	26/05/2021
								A070504	Premature Discharge of Battery	26/05/2021	26/05/2021
						F1905	Device Revision or Replacement	A0705	Battery Problem	26/05/2021	26/05/2021
								A070504	Premature Discharge of Battery	26/05/2021	26/05/2021
				CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem	26/05/2021	26/05/2021

									A070504	Premature Discharge of Battery	26/05/2021	26/05/2021
							F08	Hospitalization or Prolonged Hospitalization	A0705	Battery Problem	26/05/2021	26/05/2021
									A070504	Premature Discharge of Battery	26/05/2021	26/05/2021
							F1905	Device Revision or Replacement	A0705	Battery Problem	26/05/2021	26/05/2021
									A070504	Premature Discharge of Battery	26/05/2021	26/05/2021
					CT2985	Active	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem	26/05/2021	26/05/2021
									A070504	Premature Discharge of Battery	26/05/2021	26/05/2021
							F08	Hospitalization or Prolonged Hospitalization	A0705	Battery Problem	26/05/2021	26/05/2021
									A070504	Premature Discharge of Battery	26/05/2021	26/05/2021
							F1905	Device Revision or Replacement	A0705	Battery Problem	26/05/2021	26/05/2021
									A070504	Premature Discharge of Battery	26/05/2021	26/05/2021
					CT3001	Implantable metal	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem	26/05/2021	26/05/2021
									A070504	Premature Discharge of Battery	26/05/2021	26/05/2021
							F08	Hospitalization or Prolonged Hospitalization	A0705	Battery Problem	26/05/2021	26/05/2021
									A070504	Premature Discharge of Battery	26/05/2021	26/05/2021
							F1905	Device Revision or Replacement	A0705	Battery Problem	26/05/2021	26/05/2021
									A070504	Premature Discharge of Battery	26/05/2021	26/05/2021
					CT3002	Implantable synthetic polymer	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem	26/05/2021	26/05/2021
									A070504	Premature Discharge of Battery	26/05/2021	26/05/2021
							F08	Hospitalization or Prolonged Hospitalization	A0705	Battery Problem	26/05/2021	26/05/2021
									A070504	Premature Discharge of Battery	26/05/2021	26/05/2021
							F1905	Device Revision or Replacement	A0705	Battery Problem	26/05/2021	26/05/2021
									A070504	Premature Discharge of Battery	26/05/2021	26/05/2021
					CT3006	Implantable inorganic material	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem	26/05/2021	26/05/2021
									A070504	Premature Discharge of Battery	26/05/2021	26/05/2021
							F08	Hospitalization or Prolonged Hospitalization	A0705	Battery Problem	26/05/2021	26/05/2021
									A070504	Premature Discharge of Battery	26/05/2021	26/05/2021
							F1905	Device Revision or Replacement	A0705	Battery Problem	26/05/2021	26/05/2021
									A070504	Premature Discharge of Battery	26/05/2021	26/05/2021
					CT314	Electrically powered	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem	26/05/2021	26/05/2021
									A070504	Premature Discharge of Battery	26/05/2021	26/05/2021
							F08	Hospitalization or Prolonged Hospitalization	A0705	Battery Problem	26/05/2021	26/05/2021
									A070504	Premature Discharge of Battery	26/05/2021	26/05/2021
							F1905	Device Revision or Replacement	A0705	Battery Problem	26/05/2021	26/05/2021
									A070504	Premature Discharge of Battery	26/05/2021	26/05/2021
					CT321	Long-term surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem	26/05/2021	26/05/2021
									A070504	Premature Discharge of Battery	26/05/2021	26/05/2021
							F08	Hospitalization or Prolonged Hospitalization	A0705	Battery Problem	26/05/2021	26/05/2021
									A070504	Premature Discharge of Battery	26/05/2021	26/05/2021
							F1905	Device Revision or Replacement	A0705	Battery Problem	26/05/2021	26/05/2021
									A070504	Premature Discharge of Battery	26/05/2021	26/05/2021
					CT449	Generators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem	26/05/2021	26/05/2021
									A070504	Premature Discharge of Battery	26/05/2021	26/05/2021
							F08	Hospitalization or Prolonged Hospitalization	A0705	Battery Problem	26/05/2021	26/05/2021
									A070504	Premature Discharge of Battery	26/05/2021	26/05/2021
							F1905	Device Revision or Replacement	A0705	Battery Problem	26/05/2021	26/05/2021
									A070504	Premature Discharge of Battery	26/05/2021	26/05/2021
					CT981	Single use	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem	26/05/2021	26/05/2021
									A070504	Premature Discharge of Battery	26/05/2021	26/05/2021
							F08	Hospitalization or Prolonged Hospitalization	A0705	Battery Problem	26/05/2021	26/05/2021
									A070504	Premature Discharge of Battery	26/05/2021	26/05/2021
							F1905	Device Revision or Replacement	A0705	Battery Problem	26/05/2021	26/05/2021
									A070504	Premature Discharge of Battery	26/05/2021	26/05/2021
					CT983	Surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem	26/05/2021	26/05/2021
									A070504	Premature Discharge of Battery	26/05/2021	26/05/2021
							F08	Hospitalization or Prolonged Hospitalization	A0705	Battery Problem	26/05/2021	26/05/2021
									A070504	Premature Discharge of Battery	26/05/2021	26/05/2021
							F1905	Device Revision or Replacement	A0705	Battery Problem	26/05/2021	26/05/2021
									A070504	Premature Discharge of Battery	26/05/2021	26/05/2021
11015	2021	All other reportable incidents	35852	Single-chamber implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E2104	Electric Shock	A071302	Inappropriate/Inadequate Shock/Stimulation		
							E2403	No Clinical Signs, Symptoms or Conditions	A071302	Inappropriate/Inadequate Shock/Stimulation		
							F26	No Health Consequences or Impact	A071302	Inappropriate/Inadequate Shock/Stimulation		
11016	2021	All other reportable incidents	37265	Dual-chamber implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A070909	Oversensing		
							F11	Non-Serious Injury/ Illness/ Impairment	A070909	Oversensing		
11017	2021	All other reportable incidents	47263	Cardiac resynchronization therapy implantable pacemaker	CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A0908	Incorrect, Inadequate or Imprecise Result or Readings	30/04/2021	30/04/2021
							F08	Hospitalization or Prolonged Hospitalization	A0908	Incorrect, Inadequate or Imprecise Result or Readings	30/04/2021	30/04/2021
							F19	Surgical Intervention	A0908	Incorrect, Inadequate or Imprecise Result or Readings	30/04/2021	30/04/2021
							F1903	Device Explantation	A0908	Incorrect, Inadequate or Imprecise Result or Readings	30/04/2021	30/04/2021
					CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A0908	Incorrect, Inadequate or Imprecise Result or Readings	30/04/2021	30/04/2021
							F08	Hospitalization or Prolonged Hospitalization	A0908	Incorrect, Inadequate or Imprecise Result or Readings	30/04/2021	30/04/2021
							F19	Surgical Intervention	A0908	Incorrect, Inadequate or Imprecise Result or Readings	30/04/2021	30/04/2021
							F1903	Device Explantation	A0908	Incorrect, Inadequate or Imprecise Result or Readings	30/04/2021	30/04/2021
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A0908	Incorrect, Inadequate or Imprecise Result or Readings	30/04/2021	30/04/2021
							F08	Hospitalization or Prolonged Hospitalization	A0908	Incorrect, Inadequate or Imprecise Result or Readings	30/04/2021	30/04/2021
							F19	Surgical Intervention	A0908	Incorrect, Inadequate or Imprecise Result or Readings	30/04/2021	30/04/2021
							F1903	Device Explantation	A0908	Incorrect, Inadequate or Imprecise Result or Readings	30/04/2021	30/04/2021
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A0908	Incorrect, Inadequate or Imprecise Result or Readings	30/04/2021	30/04/2021

							F08	Hospitalization or Prolonged Hospitalization	A0908	Incorrect, Inadequate or Imprecise Result or Readings	30/04/2021	30/04/2021
							F19	Surgical Intervention	A0908	Incorrect, Inadequate or Imprecise Result or Readings	30/04/2021	30/04/2021
							F1903	Device Explantation	A0908	Incorrect, Inadequate or Imprecise Result or Readings	30/04/2021	30/04/2021
11018	2021	All other reportable incidents	47263	Cardiac resynchronization therapy implantable pacemaker	CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A0908	Incorrect, Inadequate or Imprecise Result or Readings	01/04/2021	01/04/2021
							F08	Hospitalization or Prolonged Hospitalization	A0908	Incorrect, Inadequate or Imprecise Result or Readings	01/04/2021	01/04/2021
							F12	Serious Injury/ Illness/ Impairment	A0908	Incorrect, Inadequate or Imprecise Result or Readings	01/04/2021	01/04/2021
							F1905	Device Revision or Replacement	A0908	Incorrect, Inadequate or Imprecise Result or Readings	01/04/2021	01/04/2021
					CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A0908	Incorrect, Inadequate or Imprecise Result or Readings	01/04/2021	01/04/2021
							F08	Hospitalization or Prolonged Hospitalization	A0908	Incorrect, Inadequate or Imprecise Result or Readings	01/04/2021	01/04/2021
							F12	Serious Injury/ Illness/ Impairment	A0908	Incorrect, Inadequate or Imprecise Result or Readings	01/04/2021	01/04/2021
							F1905	Device Revision or Replacement	A0908	Incorrect, Inadequate or Imprecise Result or Readings	01/04/2021	01/04/2021
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A0908	Incorrect, Inadequate or Imprecise Result or Readings	01/04/2021	01/04/2021
							F08	Hospitalization or Prolonged Hospitalization	A0908	Incorrect, Inadequate or Imprecise Result or Readings	01/04/2021	01/04/2021
							F12	Serious Injury/ Illness/ Impairment	A0908	Incorrect, Inadequate or Imprecise Result or Readings	01/04/2021	01/04/2021
							F1905	Device Revision or Replacement	A0908	Incorrect, Inadequate or Imprecise Result or Readings	01/04/2021	01/04/2021
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A0908	Incorrect, Inadequate or Imprecise Result or Readings	01/04/2021	01/04/2021
							F08	Hospitalization or Prolonged Hospitalization	A0908	Incorrect, Inadequate or Imprecise Result or Readings	01/04/2021	01/04/2021
							F12	Serious Injury/ Illness/ Impairment	A0908	Incorrect, Inadequate or Imprecise Result or Readings	01/04/2021	01/04/2021
							F1905	Device Revision or Replacement	A0908	Incorrect, Inadequate or Imprecise Result or Readings	01/04/2021	01/04/2021
11019	2021	All other reportable incidents	47267	Single-chamber implantable pacemaker, rate-responsive	CT141	Electrodes/Leads and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	04/03/2021	04/03/2021
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	04/03/2021	04/03/2021
							F12	Serious Injury/ Illness/ Impairment	A070504	Premature Discharge of Battery	04/03/2021	04/03/2021
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	04/03/2021	04/03/2021
					CT145	Cardiovascular devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	04/03/2021	04/03/2021
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	04/03/2021	04/03/2021
							F12	Serious Injury/ Illness/ Impairment	A070504	Premature Discharge of Battery	04/03/2021	04/03/2021
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	04/03/2021	04/03/2021
					CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	04/03/2021	04/03/2021
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	04/03/2021	04/03/2021
							F12	Serious Injury/ Illness/ Impairment	A070504	Premature Discharge of Battery	04/03/2021	04/03/2021
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	04/03/2021	04/03/2021
					CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	04/03/2021	04/03/2021
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	04/03/2021	04/03/2021
							F12	Serious Injury/ Illness/ Impairment	A070504	Premature Discharge of Battery	04/03/2021	04/03/2021
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	04/03/2021	04/03/2021
					CT233	Surgical	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	04/03/2021	04/03/2021
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	04/03/2021	04/03/2021
							F12	Serious Injury/ Illness/ Impairment	A070504	Premature Discharge of Battery	04/03/2021	04/03/2021
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	04/03/2021	04/03/2021
					CT2406	Implantable	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	04/03/2021	04/03/2021
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	04/03/2021	04/03/2021
							F12	Serious Injury/ Illness/ Impairment	A070504	Premature Discharge of Battery	04/03/2021	04/03/2021
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	04/03/2021	04/03/2021
					CT2407	Active implantable	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	04/03/2021	04/03/2021
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	04/03/2021	04/03/2021
							F12	Serious Injury/ Illness/ Impairment	A070504	Premature Discharge of Battery	04/03/2021	04/03/2021
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	04/03/2021	04/03/2021
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	04/03/2021	04/03/2021
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	04/03/2021	04/03/2021
							F12	Serious Injury/ Illness/ Impairment	A070504	Premature Discharge of Battery	04/03/2021	04/03/2021
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	04/03/2021	04/03/2021
					CT2727	Active implant devices/ancillaries	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	04/03/2021	04/03/2021
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	04/03/2021	04/03/2021
							F12	Serious Injury/ Illness/ Impairment	A070504	Premature Discharge of Battery	04/03/2021	04/03/2021
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	04/03/2021	04/03/2021
					CT2750	Implantable cardiac electrical stimulators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	04/03/2021	04/03/2021
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	04/03/2021	04/03/2021
							F12	Serious Injury/ Illness/ Impairment	A070504	Premature Discharge of Battery	04/03/2021	04/03/2021
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	04/03/2021	04/03/2021
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	04/03/2021	04/03/2021
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	04/03/2021	04/03/2021
							F12	Serious Injury/ Illness/ Impairment	A070504	Premature Discharge of Battery	04/03/2021	04/03/2021
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	04/03/2021	04/03/2021
					CT2985	Active	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	04/03/2021	04/03/2021
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	04/03/2021	04/03/2021
							F12	Serious Injury/ Illness/ Impairment	A070504	Premature Discharge of Battery	04/03/2021	04/03/2021
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	04/03/2021	04/03/2021
					CT3001	Implantable metal	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	04/03/2021	04/03/2021
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	04/03/2021	04/03/2021
							F12	Serious Injury/ Illness/ Impairment	A070504	Premature Discharge of Battery	04/03/2021	04/03/2021
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	04/03/2021	04/03/2021
					CT3002	Implantable synthetic polymer	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	04/03/2021	04/03/2021
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	04/03/2021	04/03/2021

							F12	Serious Injury/ Illness/ Impairment	A070504	Premature Discharge of Battery	04/03/2021	04/03/2021
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	04/03/2021	04/03/2021
					CT3006	Implantable inorganic material	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	04/03/2021	04/03/2021
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	04/03/2021	04/03/2021
							F12	Serious Injury/ Illness/ Impairment	A070504	Premature Discharge of Battery	04/03/2021	04/03/2021
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	04/03/2021	04/03/2021
					CT314	Electrically powered	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	04/03/2021	04/03/2021
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	04/03/2021	04/03/2021
							F12	Serious Injury/ Illness/ Impairment	A070504	Premature Discharge of Battery	04/03/2021	04/03/2021
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	04/03/2021	04/03/2021
					CT321	Long-term surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	04/03/2021	04/03/2021
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	04/03/2021	04/03/2021
							F12	Serious Injury/ Illness/ Impairment	A070504	Premature Discharge of Battery	04/03/2021	04/03/2021
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	04/03/2021	04/03/2021
					CT449	Generators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	04/03/2021	04/03/2021
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	04/03/2021	04/03/2021
							F12	Serious Injury/ Illness/ Impairment	A070504	Premature Discharge of Battery	04/03/2021	04/03/2021
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	04/03/2021	04/03/2021
					CT981	Single use	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	04/03/2021	04/03/2021
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	04/03/2021	04/03/2021
							F12	Serious Injury/ Illness/ Impairment	A070504	Premature Discharge of Battery	04/03/2021	04/03/2021
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	04/03/2021	04/03/2021
					CT983	Surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	04/03/2021	04/03/2021
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	04/03/2021	04/03/2021
							F12	Serious Injury/ Illness/ Impairment	A070504	Premature Discharge of Battery	04/03/2021	04/03/2021
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	04/03/2021	04/03/2021
11020	2021	All other reportable incidents	47265	Dual-chamber implantable pacemaker, rate-responsive	CT141	Electrodes/Leads and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	18/02/2021	18/02/2021
							F1901	Additional Surgery	A070504	Premature Discharge of Battery	18/02/2021	18/02/2021
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	18/02/2021	18/02/2021
					CT145	Cardiovascular devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	18/02/2021	18/02/2021
							F1901	Additional Surgery	A070504	Premature Discharge of Battery	18/02/2021	18/02/2021
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	18/02/2021	18/02/2021
					CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	18/02/2021	18/02/2021
							F1901	Additional Surgery	A070504	Premature Discharge of Battery	18/02/2021	18/02/2021
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	18/02/2021	18/02/2021
					CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	18/02/2021	18/02/2021
							F1901	Additional Surgery	A070504	Premature Discharge of Battery	18/02/2021	18/02/2021
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	18/02/2021	18/02/2021
					CT233	Surgical	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	18/02/2021	18/02/2021
							F1901	Additional Surgery	A070504	Premature Discharge of Battery	18/02/2021	18/02/2021
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	18/02/2021	18/02/2021
					CT2406	Implantable	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	18/02/2021	18/02/2021
							F1901	Additional Surgery	A070504	Premature Discharge of Battery	18/02/2021	18/02/2021
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	18/02/2021	18/02/2021
					CT2407	Active implantable	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	18/02/2021	18/02/2021
							F1901	Additional Surgery	A070504	Premature Discharge of Battery	18/02/2021	18/02/2021
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	18/02/2021	18/02/2021
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	18/02/2021	18/02/2021
							F1901	Additional Surgery	A070504	Premature Discharge of Battery	18/02/2021	18/02/2021
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	18/02/2021	18/02/2021
					CT2727	Active implant devices/ancillaries	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	18/02/2021	18/02/2021
							F1901	Additional Surgery	A070504	Premature Discharge of Battery	18/02/2021	18/02/2021
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	18/02/2021	18/02/2021
					CT2750	Implantable cardiac electrical stimulators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	18/02/2021	18/02/2021
							F1901	Additional Surgery	A070504	Premature Discharge of Battery	18/02/2021	18/02/2021
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	18/02/2021	18/02/2021
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	18/02/2021	18/02/2021
							F1901	Additional Surgery	A070504	Premature Discharge of Battery	18/02/2021	18/02/2021
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	18/02/2021	18/02/2021
					CT2985	Active	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	18/02/2021	18/02/2021
							F1901	Additional Surgery	A070504	Premature Discharge of Battery	18/02/2021	18/02/2021
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	18/02/2021	18/02/2021
					CT3001	Implantable metal	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	18/02/2021	18/02/2021
							F1901	Additional Surgery	A070504	Premature Discharge of Battery	18/02/2021	18/02/2021
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	18/02/2021	18/02/2021
					CT3002	Implantable synthetic polymer	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	18/02/2021	18/02/2021
							F1901	Additional Surgery	A070504	Premature Discharge of Battery	18/02/2021	18/02/2021
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	18/02/2021	18/02/2021
					CT3006	Implantable inorganic material	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	18/02/2021	18/02/2021
							F1901	Additional Surgery	A070504	Premature Discharge of Battery	18/02/2021	18/02/2021
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	18/02/2021	18/02/2021
					CT314	Electrically powered	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	18/02/2021	18/02/2021
							F1901	Additional Surgery	A070504	Premature Discharge of Battery	18/02/2021	18/02/2021
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	18/02/2021	18/02/2021
					CT321	Long-term surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	18/02/2021	18/02/2021
							F1901	Additional Surgery	A070504	Premature Discharge of Battery	18/02/2021	18/02/2021
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	18/02/2021	18/02/2021
					CT449	Generators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	18/02/2021	18/02/2021
							F1901	Additional Surgery	A070504	Premature Discharge of Battery	18/02/2021	18/02/2021
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	18/02/2021	18/02/2021
					CT981	Single use	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	18/02/2021	18/02/2021
							F1901	Additional Surgery	A070504	Premature Discharge of Battery	18/02/2021	18/02/2021
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	18/02/2021	18/02/2021
					CT983	Surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	18/02/2021	18/02/2021
							F1901	Additional Surgery	A070504	Premature Discharge of Battery	18/02/2021	18/02/2021
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	18/02/2021	18/02/2021
11021	2021	All other reportable incidents	36241	Pacemaker/defibrillator lead	CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A010402	Migration		
									A070909	Oversensing		

									A071302	Inappropriate/Inadequate Shock/Stimulation		
							F26	No Health Consequences or Impact	A010402	Migration		
									A070909	Oversensing		
									A071302	Inappropriate/Inadequate Shock/Stimulation		
11022	2021	All other reportable incidents	47263	Cardiac resynchronization therapy implantable pacemaker	CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem		
									A070909	Oversensing		
									A0712	Pacing Problem		
									A072202	Low Impedance		
									A1304	Telemetry Discrepancy		
							F26	No Health Consequences or Impact	A0705	Battery Problem		
									A070909	Oversensing		
									A0712	Pacing Problem		
									A072202	Low Impedance		
									A1304	Telemetry Discrepancy		
					CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem		
									A070909	Oversensing		
									A0712	Pacing Problem		
									A072202	Low Impedance		
									A1304	Telemetry Discrepancy		
							F26	No Health Consequences or Impact	A0705	Battery Problem		
									A070909	Oversensing		
									A0712	Pacing Problem		
									A072202	Low Impedance		
									A1304	Telemetry Discrepancy		
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem		
									A070909	Oversensing		
									A0712	Pacing Problem		
									A072202	Low Impedance		
									A1304	Telemetry Discrepancy		
							F26	No Health Consequences or Impact	A0705	Battery Problem		
									A070909	Oversensing		
									A0712	Pacing Problem		
									A072202	Low Impedance		
									A1304	Telemetry Discrepancy		
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem		
									A070909	Oversensing		
									A0712	Pacing Problem		
									A072202	Low Impedance		
									A1304	Telemetry Discrepancy		
							F26	No Health Consequences or Impact	A0705	Battery Problem		
									A070909	Oversensing		
									A0712	Pacing Problem		
									A072202	Low Impedance		
									A1304	Telemetry Discrepancy		
11023	2021	All other reportable incidents	35852	Single-chamber implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E2104	Electric Shock	A070909	Oversensing		
									A071302	Inappropriate/Inadequate Shock/Stimulation		
							E2403	No Clinical Signs, Symptoms or Conditions	A070909	Oversensing		
									A071302	Inappropriate/Inadequate Shock/Stimulation		
							F26	No Health Consequences or Impact	A070909	Oversensing		
									A071302	Inappropriate/Inadequate Shock/Stimulation		
11024	2021	All other reportable incidents	37265	Dual-chamber implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E2104	Electric Shock	A0712	Pacing Problem		
									A071302	Inappropriate/Inadequate Shock/Stimulation		
							F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A0712	Pacing Problem		
									A071302	Inappropriate/Inadequate Shock/Stimulation		
							F15	Recognized Device or Procedural Complication	A0712	Pacing Problem		
									A071302	Inappropriate/Inadequate Shock/Stimulation		
11025	2021	All other reportable incidents	47265	Dual-chamber implantable pacemaker, rate-responsive	CT141	Electrodes/Leads and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT145	Cardiovascular devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT233	Surgical	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT2406	Implantable	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT2407	Active implantable	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT2727	Active implant devices/ancillaries	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT2750	Implantable cardiac electrical stimulators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT2985	Active	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT3001	Implantable metal	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT3002	Implantable synthetic polymer	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT3006	Implantable inorganic material	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT314	Electrically powered	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		

					CT321	Long-term surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT449	Generators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT981	Single use	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT983	Surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
11026	2021	All other reportable incidents	47265	Dual-chamber implantable pacemaker, rate-responsive	CT141	Electrodes/Leads and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	04/05/2021	04/05/2021
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	04/05/2021	04/05/2021
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	04/05/2021	04/05/2021
					CT145	Cardiovascular devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	04/05/2021	04/05/2021
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	04/05/2021	04/05/2021
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	04/05/2021	04/05/2021
					CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	04/05/2021	04/05/2021
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	04/05/2021	04/05/2021
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	04/05/2021	04/05/2021
					CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	04/05/2021	04/05/2021
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	04/05/2021	04/05/2021
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	04/05/2021	04/05/2021
					CT233	Surgical	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	04/05/2021	04/05/2021
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	04/05/2021	04/05/2021
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	04/05/2021	04/05/2021
					CT2406	Implantable	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	04/05/2021	04/05/2021
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	04/05/2021	04/05/2021
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	04/05/2021	04/05/2021
					CT2407	Active implantable	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	04/05/2021	04/05/2021
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	04/05/2021	04/05/2021
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	04/05/2021	04/05/2021
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	04/05/2021	04/05/2021
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	04/05/2021	04/05/2021
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	04/05/2021	04/05/2021
					CT2727	Active implant devices/ancillaries	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	04/05/2021	04/05/2021
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	04/05/2021	04/05/2021
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	04/05/2021	04/05/2021
					CT2750	Implantable cardiac electrical stimulators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	04/05/2021	04/05/2021
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	04/05/2021	04/05/2021
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	04/05/2021	04/05/2021
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	04/05/2021	04/05/2021
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	04/05/2021	04/05/2021
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	04/05/2021	04/05/2021
					CT2985	Active	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	04/05/2021	04/05/2021
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	04/05/2021	04/05/2021
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	04/05/2021	04/05/2021
					CT3001	Implantable metal	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	04/05/2021	04/05/2021
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	04/05/2021	04/05/2021
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	04/05/2021	04/05/2021
					CT3002	Implantable synthetic polymer	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	04/05/2021	04/05/2021
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	04/05/2021	04/05/2021
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	04/05/2021	04/05/2021
					CT3006	Implantable inorganic material	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	04/05/2021	04/05/2021
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	04/05/2021	04/05/2021
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	04/05/2021	04/05/2021
					CT314	Electrically powered	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	04/05/2021	04/05/2021
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	04/05/2021	04/05/2021
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	04/05/2021	04/05/2021
					CT321	Long-term surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	04/05/2021	04/05/2021
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	04/05/2021	04/05/2021
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	04/05/2021	04/05/2021
					CT449	Generators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	04/05/2021	04/05/2021
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	04/05/2021	04/05/2021
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	04/05/2021	04/05/2021
					CT981	Single use	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	04/05/2021	04/05/2021
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	04/05/2021	04/05/2021
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	04/05/2021	04/05/2021
					CT983	Surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	04/05/2021	04/05/2021
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	04/05/2021	04/05/2021
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	04/05/2021	04/05/2021
11027	2021	All other reportable incidents	47265	Dual-chamber implantable pacemaker, rate-responsive	CT141	Electrodes/Leads and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT145	Cardiovascular devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT233	Surgical	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT2406	Implantable	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT2407	Active implantable	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT2727	Active implant devices/ancillaries	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT2750	Implantable cardiac electrical stimulators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		

					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT2985	Active	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT3001	Implantable metal	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT3002	Implantable synthetic polymer	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT3006	Implantable inorganic material	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT314	Electrically powered	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT321	Long-term surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT449	Generators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT981	Single use	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT983	Surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
11028	2021	All other reportable incidents	17882	Physiologic-monitoring defibrillation system	CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A07	Electrical /Electronic Property Problem		
							F27	Problem Identified During Non-Clinical Procedure	A07	Electrical /Electronic Property Problem		
11029	2021	All other reportable incidents	17882	Physiologic-monitoring defibrillation system	CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070803	Failure to Power Up		
							F27	Problem Identified During Non-Clinical Procedure	A070803	Failure to Power Up		
11030	2021	All other reportable incidents	47265	Dual-chamber implantable pacemaker, rate-responsive	CT141	Electrodes/Leads and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT145	Cardiovascular devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT233	Surgical	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT2406	Implantable	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT2407	Active implantable	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT2727	Active implant devices/ancillaries	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT2750	Implantable cardiac electrical stimulators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT2985	Active	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT3001	Implantable metal	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT3002	Implantable synthetic polymer	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT3006	Implantable inorganic material	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT314	Electrically powered	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT321	Long-term surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT449	Generators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT981	Single use	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT983	Surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
11031	2021	All other reportable incidents	47265	Dual-chamber implantable pacemaker, rate-responsive	CT141	Electrodes/Leads and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT145	Cardiovascular devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT233	Surgical	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT2406	Implantable	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT2407	Active implantable	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT2727	Active implant devices/ancillaries	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT2750	Implantable cardiac electrical stimulators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT2985	Active	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		

					CT3001	Implantable metal	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT3002	Implantable synthetic polymer	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT3006	Implantable inorganic material	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT314	Electrically powered	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT321	Long-term surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT449	Generators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT981	Single use	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT983	Surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
11032	2021	All other reportable incidents	47263	Cardiac resynchronization therapy implantable pacemaker	CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
11033	2021	All other reportable incidents	36241	Pacemaker/defibrillator lead	CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A072201	High Impedance		
							F11	Non-Serious Injury/ Illness/ Impairment	A072201	High Impedance		
11034	2021	All other reportable incidents	17882	Physiologic-monitoring defibrillation system	CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A07	Electrical /Electronic Property Problem		
							F27	Problem Identified During Non-Clinical Procedure	A07	Electrical /Electronic Property Problem		
11035	2021	All other reportable incidents	35852	Single-chamber implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A160601	Premature Elective Replacement Indicator		
							F1903	Device Explantation	A160601	Premature Elective Replacement Indicator		
							F1905	Device Revision or Replacement	A160601	Premature Elective Replacement Indicator		
11036	2021	All other reportable incidents	35852	Single-chamber implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	13/05/2021	13/05/2021
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery	13/05/2021	13/05/2021
11037	2021	All other reportable incidents	47265	Dual-chamber implantable pacemaker, rate-responsive	CT141	Electrodes/Leads and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT145	Cardiovascular devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT233	Surgical	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT2406	Implantable	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT2407	Active implantable	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT2727	Active implant devices/ancillaries	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT2750	Implantable cardiac electrical stimulators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT2985	Active	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT3001	Implantable metal	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT3002	Implantable synthetic polymer	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT3006	Implantable inorganic material	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT314	Electrically powered	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT321	Long-term surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT449	Generators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT981	Single use	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT983	Surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
11038	2021	All other reportable incidents	35852	Single-chamber implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem	25/05/2021	25/05/2021
									A070504	Premature Discharge of Battery	25/05/2021	25/05/2021
							F08	Hospitalization or Prolonged Hospitalization	A0705	Battery Problem	25/05/2021	25/05/2021
									A070504	Premature Discharge of Battery	25/05/2021	25/05/2021
							F1905	Device Revision or Replacement	A0705	Battery Problem	25/05/2021	25/05/2021
									A070504	Premature Discharge of Battery	25/05/2021	25/05/2021
11039	2021	All other reportable incidents	17882	Physiologic-monitoring defibrillation system	CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A0719	Unexpected Shutdown		
							F26	No Health Consequences or Impact	A0719	Unexpected Shutdown		
11040	2021	All other reportable incidents	36241	Pacemaker/defibrillator lead	CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A010402	Migration		
							F08	Hospitalization or Prolonged Hospitalization	A010402	Migration		
							F1904	Device Repositioning	A010402	Migration		
							F1905	Device Revision or Replacement	A010402	Migration		
11041	2021	All other reportable incidents	47265	Dual-chamber implantable pacemaker, rate-responsive	CT141	Electrodes/Leads and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	06/05/2021	06/05/2021
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	06/05/2021	06/05/2021
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	06/05/2021	06/05/2021

					CT145	Cardiovascular devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	06/05/2021	06/05/2021
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	06/05/2021	06/05/2021
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	06/05/2021	06/05/2021
					CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	06/05/2021	06/05/2021
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	06/05/2021	06/05/2021
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	06/05/2021	06/05/2021
					CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	06/05/2021	06/05/2021
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	06/05/2021	06/05/2021
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	06/05/2021	06/05/2021
					CT233	Surgical	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	06/05/2021	06/05/2021
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	06/05/2021	06/05/2021
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	06/05/2021	06/05/2021
					CT2406	Implantable	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	06/05/2021	06/05/2021
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	06/05/2021	06/05/2021
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	06/05/2021	06/05/2021
					CT2407	Active implantable	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	06/05/2021	06/05/2021
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	06/05/2021	06/05/2021
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	06/05/2021	06/05/2021
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	06/05/2021	06/05/2021
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	06/05/2021	06/05/2021
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	06/05/2021	06/05/2021
					CT2727	Active implant devices/ancillaries	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	06/05/2021	06/05/2021
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	06/05/2021	06/05/2021
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	06/05/2021	06/05/2021
					CT2750	Implantable cardiac electrical stimulators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	06/05/2021	06/05/2021
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	06/05/2021	06/05/2021
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	06/05/2021	06/05/2021
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	06/05/2021	06/05/2021
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	06/05/2021	06/05/2021
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	06/05/2021	06/05/2021
					CT2985	Active	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	06/05/2021	06/05/2021
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	06/05/2021	06/05/2021
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	06/05/2021	06/05/2021
					CT3001	Implantable metal	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	06/05/2021	06/05/2021
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	06/05/2021	06/05/2021
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	06/05/2021	06/05/2021
					CT3002	Implantable synthetic polymer	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	06/05/2021	06/05/2021
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	06/05/2021	06/05/2021
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	06/05/2021	06/05/2021
					CT3006	Implantable inorganic material	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	06/05/2021	06/05/2021
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	06/05/2021	06/05/2021
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	06/05/2021	06/05/2021
					CT314	Electrically powered	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	06/05/2021	06/05/2021
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	06/05/2021	06/05/2021
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	06/05/2021	06/05/2021
					CT321	Long-term surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	06/05/2021	06/05/2021
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	06/05/2021	06/05/2021
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	06/05/2021	06/05/2021
					CT449	Generators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	06/05/2021	06/05/2021
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	06/05/2021	06/05/2021
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	06/05/2021	06/05/2021
					CT981	Single use	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	06/05/2021	06/05/2021
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	06/05/2021	06/05/2021
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	06/05/2021	06/05/2021
					CT983	Surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	06/05/2021	06/05/2021
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	06/05/2021	06/05/2021
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	06/05/2021	06/05/2021
11042	2021	Death	36241	Pacemaker/defibrillator lead	CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A25	No Apparent Adverse Event		
							F02	Death	A25	No Apparent Adverse Event		
11043	2021	All other reportable incidents	47265	Dual-chamber implantable pacemaker, rate-responsive	CT141	Electrodes/Leads and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
					CT145	Cardiovascular devices	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
					CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
					CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
					CT233	Surgical	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
					CT2406	Implantable	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
					CT2407	Active implantable	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem		

								A070504	Premature Discharge of Battery		
						F26	No Health Consequences or Impact	A0705	Battery Problem		
								A070504	Premature Discharge of Battery		
					CT2727	Active implant devices/ancillaries	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
					CT2750	Implantable cardiac electrical stimulators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
					CT2985	Active	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
					CT3001	Implantable metal	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
					CT3002	Implantable synthetic polymer	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
					CT3006	Implantable inorganic material	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
					CT314	Electrically powered	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
					CT321	Long-term surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
					CT449	Generators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
					CT981	Single use	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
					CT983	Surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
11044	2021	All other reportable incidents	47263	Cardiac resynchronization therapy implantable pacemaker	CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A0908	Incorrect, Inadequate or Imprecise Result or Readings	11/05/2021	11/05/2021
							F08	Hospitalization or Prolonged Hospitalization	A0908	Incorrect, Inadequate or Imprecise Result or Readings	11/05/2021	11/05/2021
							F1905	Device Revision or Replacement	A0908	Incorrect, Inadequate or Imprecise Result or Readings	11/05/2021	11/05/2021
					CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A0908	Incorrect, Inadequate or Imprecise Result or Readings	11/05/2021	11/05/2021
							F08	Hospitalization or Prolonged Hospitalization	A0908	Incorrect, Inadequate or Imprecise Result or Readings	11/05/2021	11/05/2021
							F1905	Device Revision or Replacement	A0908	Incorrect, Inadequate or Imprecise Result or Readings	11/05/2021	11/05/2021
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A0908	Incorrect, Inadequate or Imprecise Result or Readings	11/05/2021	11/05/2021
							F08	Hospitalization or Prolonged Hospitalization	A0908	Incorrect, Inadequate or Imprecise Result or Readings	11/05/2021	11/05/2021
							F1905	Device Revision or Replacement	A0908	Incorrect, Inadequate or Imprecise Result or Readings	11/05/2021	11/05/2021
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A0908	Incorrect, Inadequate or Imprecise Result or Readings	11/05/2021	11/05/2021
							F08	Hospitalization or Prolonged Hospitalization	A0908	Incorrect, Inadequate or Imprecise Result or Readings	11/05/2021	11/05/2021
							F1905	Device Revision or Replacement	A0908	Incorrect, Inadequate or Imprecise Result or Readings	11/05/2021	11/05/2021
11045	2021	All other reportable incidents	47263	Cardiac resynchronization therapy implantable pacemaker	CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A0908	Incorrect, Inadequate or Imprecise Result or Readings	06/05/2021	06/05/2021
							F08	Hospitalization or Prolonged Hospitalization	A0908	Incorrect, Inadequate or Imprecise Result or Readings	06/05/2021	06/05/2021
							F1903	Device Explantation	A0908	Incorrect, Inadequate or Imprecise Result or Readings	06/05/2021	06/05/2021
							F1905	Device Revision or Replacement	A0908	Incorrect, Inadequate or Imprecise Result or Readings	06/05/2021	06/05/2021
					CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A0908	Incorrect, Inadequate or Imprecise Result or Readings	06/05/2021	06/05/2021
							F08	Hospitalization or Prolonged Hospitalization	A0908	Incorrect, Inadequate or Imprecise Result or Readings	06/05/2021	06/05/2021
							F1903	Device Explantation	A0908	Incorrect, Inadequate or Imprecise Result or Readings	06/05/2021	06/05/2021
							F1905	Device Revision or Replacement	A0908	Incorrect, Inadequate or Imprecise Result or Readings	06/05/2021	06/05/2021
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A0908	Incorrect, Inadequate or Imprecise Result or Readings	06/05/2021	06/05/2021

						F08	Hospitalization or Prolonged Hospitalization	A0908	Incorrect, Inadequate or Imprecise Result or Readings	06/05/2021	06/05/2021
						F1903	Device Explantation	A0908	Incorrect, Inadequate or Imprecise Result or Readings	06/05/2021	06/05/2021
						F1905	Device Revision or Replacement	A0908	Incorrect, Inadequate or Imprecise Result or Readings	06/05/2021	06/05/2021
				CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A0908	Incorrect, Inadequate or Imprecise Result or Readings	06/05/2021	06/05/2021
						F08	Hospitalization or Prolonged Hospitalization	A0908	Incorrect, Inadequate or Imprecise Result or Readings	06/05/2021	06/05/2021
						F1903	Device Explantation	A0908	Incorrect, Inadequate or Imprecise Result or Readings	06/05/2021	06/05/2021
						F1905	Device Revision or Replacement	A0908	Incorrect, Inadequate or Imprecise Result or Readings	06/05/2021	06/05/2021
11046	2021	All other reportable incidents	35852	Single-chamber implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E2104	Electric Shock	A070909	Oversensing		
									A071302	Inappropriate/Inadequate Shock/Stimulation		
							E2403	No Clinical Signs, Symptoms or Conditions	A070909	Oversensing		
									A071302	Inappropriate/Inadequate Shock/Stimulation		
							F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A070909	Oversensing		
									A071302	Inappropriate/Inadequate Shock/Stimulation		
11047	2021	All other reportable incidents	47265	Dual-chamber implantable pacemaker, rate-responsive	CT141	Electrodes/Leads and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem	29/09/2021	29/09/2021
									A070504	Premature Discharge of Battery	29/09/2021	29/09/2021
							F08	Hospitalization or Prolonged Hospitalization	A0705	Battery Problem	29/09/2021	29/09/2021
									A070504	Premature Discharge of Battery	29/09/2021	29/09/2021
							F1905	Device Revision or Replacement	A0705	Battery Problem	29/09/2021	29/09/2021
									A070504	Premature Discharge of Battery	29/09/2021	29/09/2021
					CT145	Cardiovascular devices	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem	29/09/2021	29/09/2021
									A070504	Premature Discharge of Battery	29/09/2021	29/09/2021
							F08	Hospitalization or Prolonged Hospitalization	A0705	Battery Problem	29/09/2021	29/09/2021
									A070504	Premature Discharge of Battery	29/09/2021	29/09/2021
							F1905	Device Revision or Replacement	A0705	Battery Problem	29/09/2021	29/09/2021
									A070504	Premature Discharge of Battery	29/09/2021	29/09/2021
					CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem	29/09/2021	29/09/2021
									A070504	Premature Discharge of Battery	29/09/2021	29/09/2021
							F08	Hospitalization or Prolonged Hospitalization	A0705	Battery Problem	29/09/2021	29/09/2021
									A070504	Premature Discharge of Battery	29/09/2021	29/09/2021
							F1905	Device Revision or Replacement	A0705	Battery Problem	29/09/2021	29/09/2021
									A070504	Premature Discharge of Battery	29/09/2021	29/09/2021
					CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem	29/09/2021	29/09/2021
									A070504	Premature Discharge of Battery	29/09/2021	29/09/2021
							F08	Hospitalization or Prolonged Hospitalization	A0705	Battery Problem	29/09/2021	29/09/2021
									A070504	Premature Discharge of Battery	29/09/2021	29/09/2021
							F1905	Device Revision or Replacement	A0705	Battery Problem	29/09/2021	29/09/2021
									A070504	Premature Discharge of Battery	29/09/2021	29/09/2021
					CT233	Surgical	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem	29/09/2021	29/09/2021
									A070504	Premature Discharge of Battery	29/09/2021	29/09/2021
							F08	Hospitalization or Prolonged Hospitalization	A0705	Battery Problem	29/09/2021	29/09/2021
									A070504	Premature Discharge of Battery	29/09/2021	29/09/2021
							F1905	Device Revision or Replacement	A0705	Battery Problem	29/09/2021	29/09/2021
									A070504	Premature Discharge of Battery	29/09/2021	29/09/2021
					CT2406	Implantable	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem	29/09/2021	29/09/2021
									A070504	Premature Discharge of Battery	29/09/2021	29/09/2021
							F08	Hospitalization or Prolonged Hospitalization	A0705	Battery Problem	29/09/2021	29/09/2021
									A070504	Premature Discharge of Battery	29/09/2021	29/09/2021
							F1905	Device Revision or Replacement	A0705	Battery Problem	29/09/2021	29/09/2021
									A070504	Premature Discharge of Battery	29/09/2021	29/09/2021
					CT2407	Active implantable	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem	29/09/2021	29/09/2021
									A070504	Premature Discharge of Battery	29/09/2021	29/09/2021
							F08	Hospitalization or Prolonged Hospitalization	A0705	Battery Problem	29/09/2021	29/09/2021
									A070504	Premature Discharge of Battery	29/09/2021	29/09/2021
							F1905	Device Revision or Replacement	A0705	Battery Problem	29/09/2021	29/09/2021
									A070504	Premature Discharge of Battery	29/09/2021	29/09/2021
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem	29/09/2021	29/09/2021
									A070504	Premature Discharge of Battery	29/09/2021	29/09/2021
							F08	Hospitalization or Prolonged Hospitalization	A0705	Battery Problem	29/09/2021	29/09/2021
									A070504	Premature Discharge of Battery	29/09/2021	29/09/2021
							F1905	Device Revision or Replacement	A0705	Battery Problem	29/09/2021	29/09/2021
									A070504	Premature Discharge of Battery	29/09/2021	29/09/2021
					CT2727	Active implant devices/ancillaries	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem	29/09/2021	29/09/2021
									A070504	Premature Discharge of Battery	29/09/2021	29/09/2021
							F08	Hospitalization or Prolonged Hospitalization	A0705	Battery Problem	29/09/2021	29/09/2021
									A070504	Premature Discharge of Battery	29/09/2021	29/09/2021
							F1905	Device Revision or Replacement	A0705	Battery Problem	29/09/2021	29/09/2021
									A070504	Premature Discharge of Battery	29/09/2021	29/09/2021
					CT2750	Implantable cardiac electrical stimulators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem	29/09/2021	29/09/2021
									A070504	Premature Discharge of Battery	29/09/2021	29/09/2021
							F08	Hospitalization or Prolonged Hospitalization	A0705	Battery Problem	29/09/2021	29/09/2021
									A070504	Premature Discharge of Battery	29/09/2021	29/09/2021
							F1905	Device Revision or Replacement	A0705	Battery Problem	29/09/2021	29/09/2021
									A070504	Premature Discharge of Battery	29/09/2021	29/09/2021
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem	29/09/2021	29/09/2021
									A070504	Premature Discharge of Battery	29/09/2021	29/09/2021
							F08	Hospitalization or Prolonged Hospitalization	A0705	Battery Problem	29/09/2021	29/09/2021
									A070504	Premature Discharge of Battery	29/09/2021	29/09/2021
							F1905	Device Revision or Replacement	A0705	Battery Problem	29/09/2021	29/09/2021
									A070504	Premature Discharge of Battery	29/09/2021	29/09/2021
					CT2985	Active	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem	29/09/2021	29/09/2021
									A070504	Premature Discharge of Battery	29/09/2021	29/09/2021
							F08	Hospitalization or Prolonged Hospitalization	A0705	Battery Problem	29/09/2021	29/09/2021

								A070504	Premature Discharge of Battery	29/09/2021	29/09/2021
						F1905	Device Revision or Replacement	A0705	Battery Problem	29/09/2021	29/09/2021
								A070504	Premature Discharge of Battery	29/09/2021	29/09/2021
					CT3001	Implantable metal	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem	29/09/2021	29/09/2021
									A070504	Premature Discharge of Battery	29/09/2021	29/09/2021
							F08	Hospitalization or Prolonged Hospitalization	A0705	Battery Problem	29/09/2021	29/09/2021
									A070504	Premature Discharge of Battery	29/09/2021	29/09/2021
							F1905	Device Revision or Replacement	A0705	Battery Problem	29/09/2021	29/09/2021
									A070504	Premature Discharge of Battery	29/09/2021	29/09/2021
					CT3002	Implantable synthetic polymer	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem	29/09/2021	29/09/2021
									A070504	Premature Discharge of Battery	29/09/2021	29/09/2021
							F08	Hospitalization or Prolonged Hospitalization	A0705	Battery Problem	29/09/2021	29/09/2021
									A070504	Premature Discharge of Battery	29/09/2021	29/09/2021
							F1905	Device Revision or Replacement	A0705	Battery Problem	29/09/2021	29/09/2021
									A070504	Premature Discharge of Battery	29/09/2021	29/09/2021
					CT3006	Implantable inorganic material	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem	29/09/2021	29/09/2021
									A070504	Premature Discharge of Battery	29/09/2021	29/09/2021
							F08	Hospitalization or Prolonged Hospitalization	A0705	Battery Problem	29/09/2021	29/09/2021
									A070504	Premature Discharge of Battery	29/09/2021	29/09/2021
							F1905	Device Revision or Replacement	A0705	Battery Problem	29/09/2021	29/09/2021
									A070504	Premature Discharge of Battery	29/09/2021	29/09/2021
					CT314	Electrically powered	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem	29/09/2021	29/09/2021
									A070504	Premature Discharge of Battery	29/09/2021	29/09/2021
							F08	Hospitalization or Prolonged Hospitalization	A0705	Battery Problem	29/09/2021	29/09/2021
									A070504	Premature Discharge of Battery	29/09/2021	29/09/2021
							F1905	Device Revision or Replacement	A0705	Battery Problem	29/09/2021	29/09/2021
									A070504	Premature Discharge of Battery	29/09/2021	29/09/2021
					CT321	Long-term surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem	29/09/2021	29/09/2021
									A070504	Premature Discharge of Battery	29/09/2021	29/09/2021
							F08	Hospitalization or Prolonged Hospitalization	A0705	Battery Problem	29/09/2021	29/09/2021
									A070504	Premature Discharge of Battery	29/09/2021	29/09/2021
							F1905	Device Revision or Replacement	A0705	Battery Problem	29/09/2021	29/09/2021
									A070504	Premature Discharge of Battery	29/09/2021	29/09/2021
					CT449	Generators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem	29/09/2021	29/09/2021
									A070504	Premature Discharge of Battery	29/09/2021	29/09/2021
							F08	Hospitalization or Prolonged Hospitalization	A0705	Battery Problem	29/09/2021	29/09/2021
									A070504	Premature Discharge of Battery	29/09/2021	29/09/2021
							F1905	Device Revision or Replacement	A0705	Battery Problem	29/09/2021	29/09/2021
									A070504	Premature Discharge of Battery	29/09/2021	29/09/2021
					CT981	Single use	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem	29/09/2021	29/09/2021
									A070504	Premature Discharge of Battery	29/09/2021	29/09/2021
							F08	Hospitalization or Prolonged Hospitalization	A0705	Battery Problem	29/09/2021	29/09/2021
									A070504	Premature Discharge of Battery	29/09/2021	29/09/2021
							F1905	Device Revision or Replacement	A0705	Battery Problem	29/09/2021	29/09/2021
									A070504	Premature Discharge of Battery	29/09/2021	29/09/2021
					CT983	Surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem	29/09/2021	29/09/2021
									A070504	Premature Discharge of Battery	29/09/2021	29/09/2021
							F08	Hospitalization or Prolonged Hospitalization	A0705	Battery Problem	29/09/2021	29/09/2021
									A070504	Premature Discharge of Battery	29/09/2021	29/09/2021
							F1905	Device Revision or Replacement	A0705	Battery Problem	29/09/2021	29/09/2021
									A070504	Premature Discharge of Battery	29/09/2021	29/09/2021
11048	2021	All other reportable incidents	47265	Dual-chamber implantable pacemaker, rate-responsive	CT141	Electrodes/Leads and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
					CT145	Cardiovascular devices	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
					CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
					CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
					CT233	Surgical	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
					CT2406	Implantable	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
					CT2407	Active implantable	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
					CT2727	Active implant devices/ancillaries	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
					CT2750	Implantable cardiac electrical stimulators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem		

									A070504	Premature Discharge of Battery			
							F26	No Health Consequences or Impact	A0705	Battery Problem			
									A070504	Premature Discharge of Battery			
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem			
									A070504	Premature Discharge of Battery			
							F26	No Health Consequences or Impact	A0705	Battery Problem			
									A070504	Premature Discharge of Battery			
					CT2985	Active	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem			
									A070504	Premature Discharge of Battery			
							F26	No Health Consequences or Impact	A0705	Battery Problem			
									A070504	Premature Discharge of Battery			
					CT3001	Implantable metal	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem			
									A070504	Premature Discharge of Battery			
							F26	No Health Consequences or Impact	A0705	Battery Problem			
									A070504	Premature Discharge of Battery			
					CT3002	Implantable synthetic polymer	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem			
									A070504	Premature Discharge of Battery			
							F26	No Health Consequences or Impact	A0705	Battery Problem			
									A070504	Premature Discharge of Battery			
					CT3006	Implantable inorganic material	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem			
									A070504	Premature Discharge of Battery			
							F26	No Health Consequences or Impact	A0705	Battery Problem			
									A070504	Premature Discharge of Battery			
					CT314	Electrically powered	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem			
									A070504	Premature Discharge of Battery			
							F26	No Health Consequences or Impact	A0705	Battery Problem			
									A070504	Premature Discharge of Battery			
					CT321	Long-term surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem			
									A070504	Premature Discharge of Battery			
							F26	No Health Consequences or Impact	A0705	Battery Problem			
									A070504	Premature Discharge of Battery			
					CT449	Generators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem			
									A070504	Premature Discharge of Battery			
							F26	No Health Consequences or Impact	A0705	Battery Problem			
									A070504	Premature Discharge of Battery			
					CT981	Single use	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem			
									A070504	Premature Discharge of Battery			
							F26	No Health Consequences or Impact	A0705	Battery Problem			
									A070504	Premature Discharge of Battery			
					CT983	Surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem			
									A070504	Premature Discharge of Battery			
							F26	No Health Consequences or Impact	A0705	Battery Problem			
									A070504	Premature Discharge of Battery			
11049	2021	All other reportable incidents	37265	Dual-chamber implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A160601	Premature Elective Replacement Indicator	30/03/2021	30/03/2021	
							F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A160601	Premature Elective Replacement Indicator	30/03/2021	30/03/2021	
							F1905	Device Revision or Replacement	A160601	Premature Elective Replacement Indicator	30/03/2021	30/03/2021	
11050	2021	All other reportable incidents	35852	Single-chamber implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A160601	Premature Elective Replacement Indicator	31/03/2021	31/03/2021	
							F1905	Device Revision or Replacement	A160601	Premature Elective Replacement Indicator	31/03/2021	31/03/2021	
11051	2021	All other reportable incidents	47270	Cardiac resynchronization therapy implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	30/04/2021	30/04/2021	
							F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A070504	Premature Discharge of Battery	30/04/2021	30/04/2021	
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	30/04/2021	30/04/2021	
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery	30/04/2021	30/04/2021	
11052	2021	All other reportable incidents	47270	Cardiac resynchronization therapy implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery			
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery			
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery			
11053	2021	All other reportable incidents	35852	Single-chamber implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	12/08/2021	12/08/2021	
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	12/08/2021	12/08/2021	
							F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A070504	Premature Discharge of Battery	12/08/2021	12/08/2021	
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	12/08/2021	12/08/2021	
11054	2021	All other reportable incidents	35852	Single-chamber implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E011903	Syncope/Fainting	A0203	Defective Device			
							F12	Serious Injury/ Illness/ Impairment	A0203	Defective Device			
							F21	Unexpected Deterioration	A0203	Defective Device			
11055	2021	All other reportable incidents	47265	Dual-chamber implantable pacemaker, rate-responsive	CT141	Electrodes/Leads and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem	19/09/2021	19/09/2021	
									A070504	Premature Discharge of Battery	19/09/2021	19/09/2021	
							F08	Hospitalization or Prolonged Hospitalization	A0705	Battery Problem	19/09/2021	19/09/2021	
									A070504	Premature Discharge of Battery	19/09/2021	19/09/2021	
							F1905	Device Revision or Replacement	A0705	Battery Problem	19/09/2021	19/09/2021	
									A070504	Premature Discharge of Battery	19/09/2021	19/09/2021	
					CT145	Cardiovascular devices	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem	19/09/2021	19/09/2021	
									A070504	Premature Discharge of Battery	19/09/2021	19/09/2021	
							F08	Hospitalization or Prolonged Hospitalization	A0705	Battery Problem	19/09/2021	19/09/2021	
									A070504	Premature Discharge of Battery	19/09/2021	19/09/2021	
							F1905	Device Revision or Replacement	A0705	Battery Problem	19/09/2021	19/09/2021	
									A070504	Premature Discharge of Battery	19/09/2021	19/09/2021	
					CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem	19/09/2021	19/09/2021	
									A070504	Premature Discharge of Battery	19/09/2021	19/09/2021	
							F08	Hospitalization or Prolonged Hospitalization	A0705	Battery Problem	19/09/2021	19/09/2021	
									A070504	Premature Discharge of Battery	19/09/2021	19/09/2021	
							F1905	Device Revision or Replacement	A0705	Battery Problem	19/09/2021	19/09/2021	
									A070504	Premature Discharge of Battery	19/09/2021	19/09/2021	
					CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem	19/09/2021	19/09/2021	
									A070504	Premature Discharge of Battery	19/09/2021	19/09/2021	
							F08	Hospitalization or Prolonged Hospitalization	A0705	Battery Problem	19/09/2021	19/09/2021	
									A070504	Premature Discharge of Battery	19/09/2021	19/09/2021	
							F1905	Device Revision or Replacement	A0705	Battery Problem	19/09/2021	19/09/2021	
									A070504	Premature Discharge of Battery	19/09/2021	19/09/2021	
					CT233	Surgical	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem	19/09/2021	19/09/2021	
									A070504	Premature Discharge of Battery	19/09/2021	19/09/2021	

		F08	Hospitalization or Prolonged Hospitalization	A0705	Battery Problem	19/09/2021	19/09/2021
				A070504	Premature Discharge of Battery	19/09/2021	19/09/2021
		F1905	Device Revision or Replacement	A0705	Battery Problem	19/09/2021	19/09/2021
				A070504	Premature Discharge of Battery	19/09/2021	19/09/2021
CT2406	Implantable	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem	19/09/2021	19/09/2021
				A070504	Premature Discharge of Battery	19/09/2021	19/09/2021
		F08	Hospitalization or Prolonged Hospitalization	A0705	Battery Problem	19/09/2021	19/09/2021
				A070504	Premature Discharge of Battery	19/09/2021	19/09/2021
		F1905	Device Revision or Replacement	A0705	Battery Problem	19/09/2021	19/09/2021
				A070504	Premature Discharge of Battery	19/09/2021	19/09/2021
CT2407	Active implantable	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem	19/09/2021	19/09/2021
				A070504	Premature Discharge of Battery	19/09/2021	19/09/2021
		F08	Hospitalization or Prolonged Hospitalization	A0705	Battery Problem	19/09/2021	19/09/2021
				A070504	Premature Discharge of Battery	19/09/2021	19/09/2021
		F1905	Device Revision or Replacement	A0705	Battery Problem	19/09/2021	19/09/2021
				A070504	Premature Discharge of Battery	19/09/2021	19/09/2021
CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem	19/09/2021	19/09/2021
				A070504	Premature Discharge of Battery	19/09/2021	19/09/2021
		F08	Hospitalization or Prolonged Hospitalization	A0705	Battery Problem	19/09/2021	19/09/2021
				A070504	Premature Discharge of Battery	19/09/2021	19/09/2021
		F1905	Device Revision or Replacement	A0705	Battery Problem	19/09/2021	19/09/2021
				A070504	Premature Discharge of Battery	19/09/2021	19/09/2021
CT2727	Active implant devices/ancillaries	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem	19/09/2021	19/09/2021
				A070504	Premature Discharge of Battery	19/09/2021	19/09/2021
		F08	Hospitalization or Prolonged Hospitalization	A0705	Battery Problem	19/09/2021	19/09/2021
				A070504	Premature Discharge of Battery	19/09/2021	19/09/2021
		F1905	Device Revision or Replacement	A0705	Battery Problem	19/09/2021	19/09/2021
				A070504	Premature Discharge of Battery	19/09/2021	19/09/2021
CT2750	Implantable cardiac electrical stimulators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem	19/09/2021	19/09/2021
				A070504	Premature Discharge of Battery	19/09/2021	19/09/2021
		F08	Hospitalization or Prolonged Hospitalization	A0705	Battery Problem	19/09/2021	19/09/2021
				A070504	Premature Discharge of Battery	19/09/2021	19/09/2021
		F1905	Device Revision or Replacement	A0705	Battery Problem	19/09/2021	19/09/2021
				A070504	Premature Discharge of Battery	19/09/2021	19/09/2021
CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem	19/09/2021	19/09/2021
				A070504	Premature Discharge of Battery	19/09/2021	19/09/2021
		F08	Hospitalization or Prolonged Hospitalization	A0705	Battery Problem	19/09/2021	19/09/2021
				A070504	Premature Discharge of Battery	19/09/2021	19/09/2021
		F1905	Device Revision or Replacement	A0705	Battery Problem	19/09/2021	19/09/2021
				A070504	Premature Discharge of Battery	19/09/2021	19/09/2021
CT2985	Active	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem	19/09/2021	19/09/2021
				A070504	Premature Discharge of Battery	19/09/2021	19/09/2021
		F08	Hospitalization or Prolonged Hospitalization	A0705	Battery Problem	19/09/2021	19/09/2021
				A070504	Premature Discharge of Battery	19/09/2021	19/09/2021
		F1905	Device Revision or Replacement	A0705	Battery Problem	19/09/2021	19/09/2021
				A070504	Premature Discharge of Battery	19/09/2021	19/09/2021
CT3001	Implantable metal	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem	19/09/2021	19/09/2021
				A070504	Premature Discharge of Battery	19/09/2021	19/09/2021
		F08	Hospitalization or Prolonged Hospitalization	A0705	Battery Problem	19/09/2021	19/09/2021
				A070504	Premature Discharge of Battery	19/09/2021	19/09/2021
		F1905	Device Revision or Replacement	A0705	Battery Problem	19/09/2021	19/09/2021
				A070504	Premature Discharge of Battery	19/09/2021	19/09/2021
CT3002	Implantable synthetic polymer	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem	19/09/2021	19/09/2021
				A070504	Premature Discharge of Battery	19/09/2021	19/09/2021
		F08	Hospitalization or Prolonged Hospitalization	A0705	Battery Problem	19/09/2021	19/09/2021
				A070504	Premature Discharge of Battery	19/09/2021	19/09/2021
		F1905	Device Revision or Replacement	A0705	Battery Problem	19/09/2021	19/09/2021
				A070504	Premature Discharge of Battery	19/09/2021	19/09/2021
CT3006	Implantable inorganic material	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem	19/09/2021	19/09/2021
				A070504	Premature Discharge of Battery	19/09/2021	19/09/2021
		F08	Hospitalization or Prolonged Hospitalization	A0705	Battery Problem	19/09/2021	19/09/2021
				A070504	Premature Discharge of Battery	19/09/2021	19/09/2021
		F1905	Device Revision or Replacement	A0705	Battery Problem	19/09/2021	19/09/2021
				A070504	Premature Discharge of Battery	19/09/2021	19/09/2021
CT314	Electrically powered	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem	19/09/2021	19/09/2021
				A070504	Premature Discharge of Battery	19/09/2021	19/09/2021
		F08	Hospitalization or Prolonged Hospitalization	A0705	Battery Problem	19/09/2021	19/09/2021
				A070504	Premature Discharge of Battery	19/09/2021	19/09/2021
		F1905	Device Revision or Replacement	A0705	Battery Problem	19/09/2021	19/09/2021
				A070504	Premature Discharge of Battery	19/09/2021	19/09/2021
CT321	Long-term surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem	19/09/2021	19/09/2021
				A070504	Premature Discharge of Battery	19/09/2021	19/09/2021
		F08	Hospitalization or Prolonged Hospitalization	A0705	Battery Problem	19/09/2021	19/09/2021
				A070504	Premature Discharge of Battery	19/09/2021	19/09/2021
		F1905	Device Revision or Replacement	A0705	Battery Problem	19/09/2021	19/09/2021
				A070504	Premature Discharge of Battery	19/09/2021	19/09/2021
CT449	Generators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem	19/09/2021	19/09/2021
				A070504	Premature Discharge of Battery	19/09/2021	19/09/2021
		F08	Hospitalization or Prolonged Hospitalization	A0705	Battery Problem	19/09/2021	19/09/2021
				A070504	Premature Discharge of Battery	19/09/2021	19/09/2021
		F1905	Device Revision or Replacement	A0705	Battery Problem	19/09/2021	19/09/2021
				A070504	Premature Discharge of Battery	19/09/2021	19/09/2021
CT981	Single use	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem	19/09/2021	19/09/2021
				A070504	Premature Discharge of Battery	19/09/2021	19/09/2021
		F08	Hospitalization or Prolonged Hospitalization	A0705	Battery Problem	19/09/2021	19/09/2021
				A070504	Premature Discharge of Battery	19/09/2021	19/09/2021
		F1905	Device Revision or Replacement	A0705	Battery Problem	19/09/2021	19/09/2021
				A070504	Premature Discharge of Battery	19/09/2021	19/09/2021

					CT983	Surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem	19/09/2021	19/09/2021
									A070504	Premature Discharge of Battery	19/09/2021	19/09/2021
							F08	Hospitalization or Prolonged Hospitalization	A0705	Battery Problem	19/09/2021	19/09/2021
									A070504	Premature Discharge of Battery	19/09/2021	19/09/2021
							F1905	Device Revision or Replacement	A0705	Battery Problem	19/09/2021	19/09/2021
									A070504	Premature Discharge of Battery	19/09/2021	19/09/2021
11056	2021	All other reportable incidents	47263	Cardiac resynchronization therapy implantable pacemaker	CT1684	Cardiac pacemakers and associated devices	E0112	Dizziness	A070909	Oversensing	30/04/2021	30/04/2021
									A0712	Pacing Problem	30/04/2021	30/04/2021
									A0908	Incorrect, Inadequate or Imprecise Result or Readings	30/04/2021	30/04/2021
							F19	Surgical Intervention	A070909	Oversensing	30/04/2021	30/04/2021
									A0712	Pacing Problem	30/04/2021	30/04/2021
									A0908	Incorrect, Inadequate or Imprecise Result or Readings	30/04/2021	30/04/2021
							F1905	Device Revision or Replacement	A070909	Oversensing	30/04/2021	30/04/2021
									A0712	Pacing Problem	30/04/2021	30/04/2021
									A0908	Incorrect, Inadequate or Imprecise Result or Readings	30/04/2021	30/04/2021
					CT1685	Cardiac pacemakers	E0112	Dizziness	A070909	Oversensing	30/04/2021	30/04/2021
									A0712	Pacing Problem	30/04/2021	30/04/2021
									A0908	Incorrect, Inadequate or Imprecise Result or Readings	30/04/2021	30/04/2021
							F19	Surgical Intervention	A070909	Oversensing	30/04/2021	30/04/2021
									A0712	Pacing Problem	30/04/2021	30/04/2021
									A0908	Incorrect, Inadequate or Imprecise Result or Readings	30/04/2021	30/04/2021
							F1905	Device Revision or Replacement	A070909	Oversensing	30/04/2021	30/04/2021
									A0712	Pacing Problem	30/04/2021	30/04/2021
									A0908	Incorrect, Inadequate or Imprecise Result or Readings	30/04/2021	30/04/2021
					CT257	Implantable cardiac pacemakers	E0112	Dizziness	A070909	Oversensing	30/04/2021	30/04/2021
									A0712	Pacing Problem	30/04/2021	30/04/2021
									A0908	Incorrect, Inadequate or Imprecise Result or Readings	30/04/2021	30/04/2021
							F19	Surgical Intervention	A070909	Oversensing	30/04/2021	30/04/2021
									A0712	Pacing Problem	30/04/2021	30/04/2021
									A0908	Incorrect, Inadequate or Imprecise Result or Readings	30/04/2021	30/04/2021
							F1905	Device Revision or Replacement	A070909	Oversensing	30/04/2021	30/04/2021
									A0712	Pacing Problem	30/04/2021	30/04/2021
									A0908	Incorrect, Inadequate or Imprecise Result or Readings	30/04/2021	30/04/2021
					CT2866	Implantable cardiac electrical stimulators	E0112	Dizziness	A070909	Oversensing	30/04/2021	30/04/2021
									A0712	Pacing Problem	30/04/2021	30/04/2021
									A0908	Incorrect, Inadequate or Imprecise Result or Readings	30/04/2021	30/04/2021
							F19	Surgical Intervention	A070909	Oversensing	30/04/2021	30/04/2021
									A0712	Pacing Problem	30/04/2021	30/04/2021
									A0908	Incorrect, Inadequate or Imprecise Result or Readings	30/04/2021	30/04/2021
							F1905	Device Revision or Replacement	A070909	Oversensing	30/04/2021	30/04/2021
									A0712	Pacing Problem	30/04/2021	30/04/2021
									A0908	Incorrect, Inadequate or Imprecise Result or Readings	30/04/2021	30/04/2021
11057	2021	All other reportable incidents	35852	Single-chamber implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A070909	Oversensing		
							F26	No Health Consequences or Impact	A070909	Oversensing		
11058	2021	All other reportable incidents	37265	Dual-chamber implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A0508	Noise, Audible		
									A072201	High Impedance		
							F26	No Health Consequences or Impact	A0508	Noise, Audible		
									A072201	High Impedance		
11059	2021	All other reportable incidents	35852	Single-chamber implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem	26/05/2021	26/05/2021
									A070504	Premature Discharge of Battery	26/05/2021	26/05/2021
							F1905	Device Revision or Replacement	A0705	Battery Problem	26/05/2021	26/05/2021
									A070504	Premature Discharge of Battery	26/05/2021	26/05/2021
11060	2021	All other reportable incidents	47267	Single-chamber implantable pacemaker, rate-responsive	CT141	Electrodes/Leads and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
					CT145	Cardiovascular devices	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
					CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
					CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
					CT233	Surgical	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
					CT2406	Implantable	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
					CT2407	Active implantable	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A0705	Battery Problem		

									A070504	Premature Discharge of Battery		
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
					CT2727	Active implant devices/ancillaries	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
					CT2750	Implantable cardiac electrical stimulators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
					CT2985	Active	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
					CT3001	Implantable metal	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
					CT3002	Implantable synthetic polymer	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
					CT3006	Implantable inorganic material	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
					CT314	Electrically powered	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
					CT321	Long-term surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
					CT449	Generators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
					CT981	Single use	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
					CT983	Surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
11061	2021	All other reportable incidents	47263	Cardiac resynchronization therapy implantable pacemaker	CT1684	Cardiac pacemakers and associated devices	E011902	Presyncope	A070909	Oversensing	04/05/2021	04/05/2021
									A0712	Pacing Problem	04/05/2021	04/05/2021
									A0908	Incorrect, Inadequate or Imprecise Result or Readings	04/05/2021	04/05/2021
							E0602	Cardiac Arrest	A070909	Oversensing	04/05/2021	04/05/2021
									A0712	Pacing Problem	04/05/2021	04/05/2021
									A0908	Incorrect, Inadequate or Imprecise Result or Readings	04/05/2021	04/05/2021
							F08	Hospitalization or Prolonged Hospitalization	A070909	Oversensing	04/05/2021	04/05/2021
									A0712	Pacing Problem	04/05/2021	04/05/2021
									A0908	Incorrect, Inadequate or Imprecise Result or Readings	04/05/2021	04/05/2021
							F12	Serious Injury/ Illness/ Impairment	A070909	Oversensing	04/05/2021	04/05/2021
									A0712	Pacing Problem	04/05/2021	04/05/2021
									A0908	Incorrect, Inadequate or Imprecise Result or Readings	04/05/2021	04/05/2021
							F1905	Device Revision or Replacement	A070909	Oversensing	04/05/2021	04/05/2021
									A0712	Pacing Problem	04/05/2021	04/05/2021
									A0908	Incorrect, Inadequate or Imprecise Result or Readings	04/05/2021	04/05/2021
							F21	Unexpected Deterioration	A070909	Oversensing	04/05/2021	04/05/2021
									A0712	Pacing Problem	04/05/2021	04/05/2021
									A0908	Incorrect, Inadequate or Imprecise Result or Readings	04/05/2021	04/05/2021
					CT1685	Cardiac pacemakers	E011902	Presyncope	A070909	Oversensing	04/05/2021	04/05/2021
									A0712	Pacing Problem	04/05/2021	04/05/2021
									A0908	Incorrect, Inadequate or Imprecise Result or Readings	04/05/2021	04/05/2021
							E0602	Cardiac Arrest	A070909	Oversensing	04/05/2021	04/05/2021
									A0712	Pacing Problem	04/05/2021	04/05/2021
									A0908	Incorrect, Inadequate or Imprecise Result or Readings	04/05/2021	04/05/2021
							F08	Hospitalization or Prolonged Hospitalization	A070909	Oversensing	04/05/2021	04/05/2021
									A0712	Pacing Problem	04/05/2021	04/05/2021
									A0908	Incorrect, Inadequate or Imprecise Result or Readings	04/05/2021	04/05/2021
							F12	Serious Injury/ Illness/ Impairment	A070909	Oversensing	04/05/2021	04/05/2021

									A0712	Pacing Problem	04/05/2021	04/05/2021
									A0908	Incorrect, Inadequate or Imprecise Result or Readings	04/05/2021	04/05/2021
							F1905	Device Revision or Replacement	A070909	Oversensing	04/05/2021	04/05/2021
									A0712	Pacing Problem	04/05/2021	04/05/2021
									A0908	Incorrect, Inadequate or Imprecise Result or Readings	04/05/2021	04/05/2021
							F21	Unexpected Deterioration	A070909	Oversensing	04/05/2021	04/05/2021
									A0712	Pacing Problem	04/05/2021	04/05/2021
									A0908	Incorrect, Inadequate or Imprecise Result or Readings	04/05/2021	04/05/2021
					CT257	Implantable cardiac pacemakers	E011902	Presyncope	A070909	Oversensing	04/05/2021	04/05/2021
									A0712	Pacing Problem	04/05/2021	04/05/2021
									A0908	Incorrect, Inadequate or Imprecise Result or Readings	04/05/2021	04/05/2021
							E0602	Cardiac Arrest	A070909	Oversensing	04/05/2021	04/05/2021
									A0712	Pacing Problem	04/05/2021	04/05/2021
									A0908	Incorrect, Inadequate or Imprecise Result or Readings	04/05/2021	04/05/2021
							F08	Hospitalization or Prolonged Hospitalization	A070909	Oversensing	04/05/2021	04/05/2021
									A0712	Pacing Problem	04/05/2021	04/05/2021
									A0908	Incorrect, Inadequate or Imprecise Result or Readings	04/05/2021	04/05/2021
							F12	Serious Injury/ Illness/ Impairment	A070909	Oversensing	04/05/2021	04/05/2021
									A0712	Pacing Problem	04/05/2021	04/05/2021
									A0908	Incorrect, Inadequate or Imprecise Result or Readings	04/05/2021	04/05/2021
							F1905	Device Revision or Replacement	A070909	Oversensing	04/05/2021	04/05/2021
									A0712	Pacing Problem	04/05/2021	04/05/2021
									A0908	Incorrect, Inadequate or Imprecise Result or Readings	04/05/2021	04/05/2021
							F21	Unexpected Deterioration	A070909	Oversensing	04/05/2021	04/05/2021
									A0712	Pacing Problem	04/05/2021	04/05/2021
									A0908	Incorrect, Inadequate or Imprecise Result or Readings	04/05/2021	04/05/2021
					CT2866	Implantable cardiac electrical stimulators	E011902	Presyncope	A070909	Oversensing	04/05/2021	04/05/2021
									A0712	Pacing Problem	04/05/2021	04/05/2021
									A0908	Incorrect, Inadequate or Imprecise Result or Readings	04/05/2021	04/05/2021
							E0602	Cardiac Arrest	A070909	Oversensing	04/05/2021	04/05/2021
									A0712	Pacing Problem	04/05/2021	04/05/2021
									A0908	Incorrect, Inadequate or Imprecise Result or Readings	04/05/2021	04/05/2021
							F08	Hospitalization or Prolonged Hospitalization	A070909	Oversensing	04/05/2021	04/05/2021
									A0712	Pacing Problem	04/05/2021	04/05/2021
									A0908	Incorrect, Inadequate or Imprecise Result or Readings	04/05/2021	04/05/2021
							F12	Serious Injury/ Illness/ Impairment	A070909	Oversensing	04/05/2021	04/05/2021
									A0712	Pacing Problem	04/05/2021	04/05/2021
									A0908	Incorrect, Inadequate or Imprecise Result or Readings	04/05/2021	04/05/2021
							F1905	Device Revision or Replacement	A070909	Oversensing	04/05/2021	04/05/2021
									A0712	Pacing Problem	04/05/2021	04/05/2021
									A0908	Incorrect, Inadequate or Imprecise Result or Readings	04/05/2021	04/05/2021
							F21	Unexpected Deterioration	A070909	Oversensing	04/05/2021	04/05/2021
									A0712	Pacing Problem	04/05/2021	04/05/2021
									A0908	Incorrect, Inadequate or Imprecise Result or Readings	04/05/2021	04/05/2021
11062	2021	All other reportable incidents	35852	Single-chamber implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A072201	High Impedance		
							F11	Non-Serious Injury/ Illness/ Impairment	A072201	High Impedance		
11063	2021	All other reportable incidents	36241	Pacemaker/defibrillator lead	CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A072201	High Impedance		
							F26	No Health Consequences or Impact	A072201	High Impedance		
11064	2021	All other reportable incidents	47270	Cardiac resynchronization therapy implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E2104	Electric Shock	A0712	Pacing Problem		
									A071302	Inappropriate/Inadequate Shock/Stimulation		
							F11	Non-Serious Injury/ Illness/ Impairment	A0712	Pacing Problem		
									A071302	Inappropriate/Inadequate Shock/Stimulation		
11065	2021	All other reportable incidents	35852	Single-chamber implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A070909	Oversensing	01/04/2022	01/04/2022
							F1903	Device Explantation	A070909	Oversensing	01/04/2022	01/04/2022
							F26	No Health Consequences or Impact	A070909	Oversensing	01/04/2022	01/04/2022
11066	2021	All other reportable incidents	35852	Single-chamber implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E2104	Electric Shock	A071302	Inappropriate/Inadequate Shock/Stimulation		
							E2401	Insufficient Information	A071302	Inappropriate/Inadequate Shock/Stimulation		
							F26	No Health Consequences or Impact	A071302	Inappropriate/Inadequate Shock/Stimulation		
11067	2021	All other reportable incidents	17882	Physiologic-monitoring defibrillation system	CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A071301	Failure to Deliver Shock/Stimulation		
							F26	No Health Consequences or Impact	A071301	Failure to Deliver Shock/Stimulation		
11068	2021	All other reportable incidents	35852	Single-chamber implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem	27/05/2021	27/05/2021
									A070504	Premature Discharge of Battery	27/05/2021	27/05/2021
							F08	Hospitalization or Prolonged Hospitalization	A0705	Battery Problem	27/05/2021	27/05/2021
									A070504	Premature Discharge of Battery	27/05/2021	27/05/2021
							F1905	Device Revision or Replacement	A0705	Battery Problem	27/05/2021	27/05/2021
									A070504	Premature Discharge of Battery	27/05/2021	27/05/2021
11069	2021	All other reportable incidents	35852	Single-chamber implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem	15/07/2021	15/07/2021
									A070504	Premature Discharge of Battery	15/07/2021	15/07/2021
							F08	Hospitalization or Prolonged Hospitalization	A0705	Battery Problem	15/07/2021	15/07/2021
									A070504	Premature Discharge of Battery	15/07/2021	15/07/2021
							F1905	Device Revision or Replacement	A0705	Battery Problem	15/07/2021	15/07/2021
									A070504	Premature Discharge of Battery	15/07/2021	15/07/2021
11070	2021	All other reportable incidents	37265	Dual-chamber implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E2104	Electric Shock	A070504	Premature Discharge of Battery	30/04/2021	30/04/2021
									A071302	Inappropriate/Inadequate Shock/Stimulation	30/04/2021	30/04/2021
									A0908	Incorrect, Inadequate or Imprecise Result or Readings	30/04/2021	30/04/2021
							F19	Surgical Intervention	A070504	Premature Discharge of Battery	30/04/2021	30/04/2021

									A071302	Inappropriate/Inadequate Shock/Stimulation	30/04/2021	30/04/2021
									A0908	Incorrect, Inadequate or Imprecise Result or Readings	30/04/2021	30/04/2021
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	30/04/2021	30/04/2021
									A071302	Inappropriate/Inadequate Shock/Stimulation	30/04/2021	30/04/2021
									A0908	Incorrect, Inadequate or Imprecise Result or Readings	30/04/2021	30/04/2021
11071	2021	All other reportable incidents	47267	Single-chamber implantable pacemaker, rate-responsive	CT141	Electrodes/Leads and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	07/05/2021	07/05/2021
									A071102	Failure to Interrogate	07/05/2021	07/05/2021
							F1901	Additional Surgery	A070504	Premature Discharge of Battery	07/05/2021	07/05/2021
									A071102	Failure to Interrogate	07/05/2021	07/05/2021
					CT145	Cardiovascular devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	07/05/2021	07/05/2021
									A071102	Failure to Interrogate	07/05/2021	07/05/2021
							F1901	Additional Surgery	A070504	Premature Discharge of Battery	07/05/2021	07/05/2021
									A071102	Failure to Interrogate	07/05/2021	07/05/2021
					CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	07/05/2021	07/05/2021
									A071102	Failure to Interrogate	07/05/2021	07/05/2021
							F1901	Additional Surgery	A070504	Premature Discharge of Battery	07/05/2021	07/05/2021
									A071102	Failure to Interrogate	07/05/2021	07/05/2021
					CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	07/05/2021	07/05/2021
									A071102	Failure to Interrogate	07/05/2021	07/05/2021
							F1901	Additional Surgery	A070504	Premature Discharge of Battery	07/05/2021	07/05/2021
									A071102	Failure to Interrogate	07/05/2021	07/05/2021
					CT233	Surgical	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	07/05/2021	07/05/2021
									A071102	Failure to Interrogate	07/05/2021	07/05/2021
							F1901	Additional Surgery	A070504	Premature Discharge of Battery	07/05/2021	07/05/2021
									A071102	Failure to Interrogate	07/05/2021	07/05/2021
					CT2406	Implantable	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	07/05/2021	07/05/2021
									A071102	Failure to Interrogate	07/05/2021	07/05/2021
							F1901	Additional Surgery	A070504	Premature Discharge of Battery	07/05/2021	07/05/2021
									A071102	Failure to Interrogate	07/05/2021	07/05/2021
					CT2407	Active implantable	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	07/05/2021	07/05/2021
									A071102	Failure to Interrogate	07/05/2021	07/05/2021
							F1901	Additional Surgery	A070504	Premature Discharge of Battery	07/05/2021	07/05/2021
									A071102	Failure to Interrogate	07/05/2021	07/05/2021
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	07/05/2021	07/05/2021
									A071102	Failure to Interrogate	07/05/2021	07/05/2021
							F1901	Additional Surgery	A070504	Premature Discharge of Battery	07/05/2021	07/05/2021
									A071102	Failure to Interrogate	07/05/2021	07/05/2021
					CT2727	Active implant devices/ancillaries	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	07/05/2021	07/05/2021
									A071102	Failure to Interrogate	07/05/2021	07/05/2021
							F1901	Additional Surgery	A070504	Premature Discharge of Battery	07/05/2021	07/05/2021
									A071102	Failure to Interrogate	07/05/2021	07/05/2021
					CT2750	Implantable cardiac electrical stimulators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	07/05/2021	07/05/2021
									A071102	Failure to Interrogate	07/05/2021	07/05/2021
							F1901	Additional Surgery	A070504	Premature Discharge of Battery	07/05/2021	07/05/2021
									A071102	Failure to Interrogate	07/05/2021	07/05/2021
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	07/05/2021	07/05/2021
									A071102	Failure to Interrogate	07/05/2021	07/05/2021
							F1901	Additional Surgery	A070504	Premature Discharge of Battery	07/05/2021	07/05/2021
									A071102	Failure to Interrogate	07/05/2021	07/05/2021
					CT2985	Active	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	07/05/2021	07/05/2021
									A071102	Failure to Interrogate	07/05/2021	07/05/2021
							F1901	Additional Surgery	A070504	Premature Discharge of Battery	07/05/2021	07/05/2021
									A071102	Failure to Interrogate	07/05/2021	07/05/2021
					CT3001	Implantable metal	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	07/05/2021	07/05/2021
									A071102	Failure to Interrogate	07/05/2021	07/05/2021
							F1901	Additional Surgery	A070504	Premature Discharge of Battery	07/05/2021	07/05/2021
									A071102	Failure to Interrogate	07/05/2021	07/05/2021
					CT3002	Implantable synthetic polymer	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	07/05/2021	07/05/2021
									A071102	Failure to Interrogate	07/05/2021	07/05/2021
							F1901	Additional Surgery	A070504	Premature Discharge of Battery	07/05/2021	07/05/2021
									A071102	Failure to Interrogate	07/05/2021	07/05/2021
					CT3006	Implantable inorganic material	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	07/05/2021	07/05/2021
									A071102	Failure to Interrogate	07/05/2021	07/05/2021
							F1901	Additional Surgery	A070504	Premature Discharge of Battery	07/05/2021	07/05/2021
									A071102	Failure to Interrogate	07/05/2021	07/05/2021
					CT314	Electrically powered	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	07/05/2021	07/05/2021
									A071102	Failure to Interrogate	07/05/2021	07/05/2021
							F1901	Additional Surgery	A070504	Premature Discharge of Battery	07/05/2021	07/05/2021
									A071102	Failure to Interrogate	07/05/2021	07/05/2021
					CT321	Long-term surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	07/05/2021	07/05/2021
									A071102	Failure to Interrogate	07/05/2021	07/05/2021
							F1901	Additional Surgery	A070504	Premature Discharge of Battery	07/05/2021	07/05/2021
									A071102	Failure to Interrogate	07/05/2021	07/05/2021
					CT449	Generators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	07/05/2021	07/05/2021
									A071102	Failure to Interrogate	07/05/2021	07/05/2021
							F1901	Additional Surgery	A070504	Premature Discharge of Battery	07/05/2021	07/05/2021
									A071102	Failure to Interrogate	07/05/2021	07/05/2021
					CT981	Single use	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	07/05/2021	07/05/2021
									A071102	Failure to Interrogate	07/05/2021	07/05/2021
							F1901	Additional Surgery	A070504	Premature Discharge of Battery	07/05/2021	07/05/2021
									A071102	Failure to Interrogate	07/05/2021	07/05/2021
					CT983	Surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	07/05/2021	07/05/2021
									A071102	Failure to Interrogate	07/05/2021	07/05/2021
							F1901	Additional Surgery	A070504	Premature Discharge of Battery	07/05/2021	07/05/2021
									A071102	Failure to Interrogate	07/05/2021	07/05/2021
11072	2021	All other reportable incidents	47265	Dual-chamber implantable pacemaker, rate-	CT141	Electrodes/Leads and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	13/05/2021	13/05/2021

				responsive			F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery	13/05/2021	13/05/2021
					CT145	Cardiovascular devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	13/05/2021	13/05/2021
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery	13/05/2021	13/05/2021
					CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	13/05/2021	13/05/2021
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery	13/05/2021	13/05/2021
					CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	13/05/2021	13/05/2021
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery	13/05/2021	13/05/2021
					CT233	Surgical	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	13/05/2021	13/05/2021
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery	13/05/2021	13/05/2021
					CT2406	Implantable	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	13/05/2021	13/05/2021
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery	13/05/2021	13/05/2021
					CT2407	Active implantable	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	13/05/2021	13/05/2021
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery	13/05/2021	13/05/2021
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	13/05/2021	13/05/2021
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery	13/05/2021	13/05/2021
					CT2727	Active implant devices/ancillaries	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	13/05/2021	13/05/2021
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery	13/05/2021	13/05/2021
					CT2750	Implantable cardiac electrical stimulators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	13/05/2021	13/05/2021
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery	13/05/2021	13/05/2021
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	13/05/2021	13/05/2021
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery	13/05/2021	13/05/2021
					CT2985	Active	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	13/05/2021	13/05/2021
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery	13/05/2021	13/05/2021
					CT3001	Implantable metal	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	13/05/2021	13/05/2021
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery	13/05/2021	13/05/2021
					CT3002	Implantable synthetic polymer	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	13/05/2021	13/05/2021
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery	13/05/2021	13/05/2021
					CT3006	Implantable inorganic material	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	13/05/2021	13/05/2021
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery	13/05/2021	13/05/2021
					CT314	Electrically powered	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	13/05/2021	13/05/2021
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery	13/05/2021	13/05/2021
					CT321	Long-term surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	13/05/2021	13/05/2021
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery	13/05/2021	13/05/2021
					CT449	Generators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	13/05/2021	13/05/2021
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery	13/05/2021	13/05/2021
					CT981	Single use	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	13/05/2021	13/05/2021
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery	13/05/2021	13/05/2021
					CT983	Surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	13/05/2021	13/05/2021
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery	13/05/2021	13/05/2021
11073	2021	All other reportable incidents	47265	Dual-chamber implantable pacemaker, rate-responsive	CT141	Electrodes/Leads and associated devices	E0112	Dizziness	A070909	Oversensing	30/04/2021	30/04/2021
									A0712	Pacing Problem	30/04/2021	30/04/2021
							E060101	Asystole	A070909	Oversensing	30/04/2021	30/04/2021
									A0712	Pacing Problem	30/04/2021	30/04/2021
							E0602	Cardiac Arrest	A070909	Oversensing	30/04/2021	30/04/2021
									A0712	Pacing Problem	30/04/2021	30/04/2021
							F08	Hospitalization or Prolonged Hospitalization	A070909	Oversensing	30/04/2021	30/04/2021
									A0712	Pacing Problem	30/04/2021	30/04/2021
							F1905	Device Revision or Replacement	A070909	Oversensing	30/04/2021	30/04/2021
									A0712	Pacing Problem	30/04/2021	30/04/2021
					CT145	Cardiovascular devices	E0112	Dizziness	A070909	Oversensing	30/04/2021	30/04/2021
									A0712	Pacing Problem	30/04/2021	30/04/2021
							E060101	Asystole	A070909	Oversensing	30/04/2021	30/04/2021
									A0712	Pacing Problem	30/04/2021	30/04/2021
							E0602	Cardiac Arrest	A070909	Oversensing	30/04/2021	30/04/2021
									A0712	Pacing Problem	30/04/2021	30/04/2021
							F08	Hospitalization or Prolonged Hospitalization	A070909	Oversensing	30/04/2021	30/04/2021
									A0712	Pacing Problem	30/04/2021	30/04/2021
							F1905	Device Revision or Replacement	A070909	Oversensing	30/04/2021	30/04/2021
									A0712	Pacing Problem	30/04/2021	30/04/2021
					CT1684	Cardiac pacemakers and associated devices	E0112	Dizziness	A070909	Oversensing	30/04/2021	30/04/2021
									A0712	Pacing Problem	30/04/2021	30/04/2021
							E060101	Asystole	A070909	Oversensing	30/04/2021	30/04/2021
									A0712	Pacing Problem	30/04/2021	30/04/2021
							E0602	Cardiac Arrest	A070909	Oversensing	30/04/2021	30/04/2021
									A0712	Pacing Problem	30/04/2021	30/04/2021
							F08	Hospitalization or Prolonged Hospitalization	A070909	Oversensing	30/04/2021	30/04/2021
									A0712	Pacing Problem	30/04/2021	30/04/2021
							F1905	Device Revision or Replacement	A070909	Oversensing	30/04/2021	30/04/2021
									A0712	Pacing Problem	30/04/2021	30/04/2021
					CT1685	Cardiac pacemakers	E0112	Dizziness	A070909	Oversensing	30/04/2021	30/04/2021
									A0712	Pacing Problem	30/04/2021	30/04/2021
							E060101	Asystole	A070909	Oversensing	30/04/2021	30/04/2021
									A0712	Pacing Problem	30/04/2021	30/04/2021
							E0602	Cardiac Arrest	A070909	Oversensing	30/04/2021	30/04/2021
									A0712	Pacing Problem	30/04/2021	30/04/2021
							F08	Hospitalization or Prolonged Hospitalization	A070909	Oversensing	30/04/2021	30/04/2021
									A0712	Pacing Problem	30/04/2021	30/04/2021
							F1905	Device Revision or Replacement	A070909	Oversensing	30/04/2021	30/04/2021
									A0712	Pacing Problem	30/04/2021	30/04/2021
					CT233	Surgical	E0112	Dizziness	A070909	Oversensing	30/04/2021	30/04/2021
									A0712	Pacing Problem	30/04/2021	30/04/2021
							E060101	Asystole	A070909	Oversensing	30/04/2021	30/04/2021
									A0712	Pacing Problem	30/04/2021	30/04/2021
							E0602	Cardiac Arrest	A070909	Oversensing	30/04/2021	30/04/2021
									A0712	Pacing Problem	30/04/2021	30/04/2021
							F08	Hospitalization or Prolonged Hospitalization	A070909	Oversensing	30/04/2021	30/04/2021
									A0712	Pacing Problem	30/04/2021	30/04/2021
							F1905	Device Revision or Replacement	A070909	Oversensing	30/04/2021	30/04/2021

									A0712	Pacing Problem	30/04/2021	30/04/2021
					CT2406	Implantable	E0112	Dizziness	A070909	Oversensing	30/04/2021	30/04/2021
									A0712	Pacing Problem	30/04/2021	30/04/2021
							E060101	Asystole	A070909	Oversensing	30/04/2021	30/04/2021
									A0712	Pacing Problem	30/04/2021	30/04/2021
							E0602	Cardiac Arrest	A070909	Oversensing	30/04/2021	30/04/2021
									A0712	Pacing Problem	30/04/2021	30/04/2021
							F08	Hospitalization or Prolonged Hospitalization	A070909	Oversensing	30/04/2021	30/04/2021
									A0712	Pacing Problem	30/04/2021	30/04/2021
							F1905	Device Revision or Replacement	A070909	Oversensing	30/04/2021	30/04/2021
									A0712	Pacing Problem	30/04/2021	30/04/2021
					CT2407	Active implantable	E0112	Dizziness	A070909	Oversensing	30/04/2021	30/04/2021
									A0712	Pacing Problem	30/04/2021	30/04/2021
							E060101	Asystole	A070909	Oversensing	30/04/2021	30/04/2021
									A0712	Pacing Problem	30/04/2021	30/04/2021
							E0602	Cardiac Arrest	A070909	Oversensing	30/04/2021	30/04/2021
									A0712	Pacing Problem	30/04/2021	30/04/2021
							F08	Hospitalization or Prolonged Hospitalization	A070909	Oversensing	30/04/2021	30/04/2021
									A0712	Pacing Problem	30/04/2021	30/04/2021
							F1905	Device Revision or Replacement	A070909	Oversensing	30/04/2021	30/04/2021
									A0712	Pacing Problem	30/04/2021	30/04/2021
					CT257	Implantable cardiac pacemakers	E0112	Dizziness	A070909	Oversensing	30/04/2021	30/04/2021
									A0712	Pacing Problem	30/04/2021	30/04/2021
							E060101	Asystole	A070909	Oversensing	30/04/2021	30/04/2021
									A0712	Pacing Problem	30/04/2021	30/04/2021
							E0602	Cardiac Arrest	A070909	Oversensing	30/04/2021	30/04/2021
									A0712	Pacing Problem	30/04/2021	30/04/2021
							F08	Hospitalization or Prolonged Hospitalization	A070909	Oversensing	30/04/2021	30/04/2021
									A0712	Pacing Problem	30/04/2021	30/04/2021
							F1905	Device Revision or Replacement	A070909	Oversensing	30/04/2021	30/04/2021
									A0712	Pacing Problem	30/04/2021	30/04/2021
					CT2727	Active implant devices/ancillaries	E0112	Dizziness	A070909	Oversensing	30/04/2021	30/04/2021
									A0712	Pacing Problem	30/04/2021	30/04/2021
							E060101	Asystole	A070909	Oversensing	30/04/2021	30/04/2021
									A0712	Pacing Problem	30/04/2021	30/04/2021
							E0602	Cardiac Arrest	A070909	Oversensing	30/04/2021	30/04/2021
									A0712	Pacing Problem	30/04/2021	30/04/2021
							F08	Hospitalization or Prolonged Hospitalization	A070909	Oversensing	30/04/2021	30/04/2021
									A0712	Pacing Problem	30/04/2021	30/04/2021
							F1905	Device Revision or Replacement	A070909	Oversensing	30/04/2021	30/04/2021
									A0712	Pacing Problem	30/04/2021	30/04/2021
					CT2750	Implantable cardiac electrical stimulators and associated devices	E0112	Dizziness	A070909	Oversensing	30/04/2021	30/04/2021
									A0712	Pacing Problem	30/04/2021	30/04/2021
							E060101	Asystole	A070909	Oversensing	30/04/2021	30/04/2021
									A0712	Pacing Problem	30/04/2021	30/04/2021
							E0602	Cardiac Arrest	A070909	Oversensing	30/04/2021	30/04/2021
									A0712	Pacing Problem	30/04/2021	30/04/2021
							F08	Hospitalization or Prolonged Hospitalization	A070909	Oversensing	30/04/2021	30/04/2021
									A0712	Pacing Problem	30/04/2021	30/04/2021
							F1905	Device Revision or Replacement	A070909	Oversensing	30/04/2021	30/04/2021
									A0712	Pacing Problem	30/04/2021	30/04/2021
					CT2866	Implantable cardiac electrical stimulators	E0112	Dizziness	A070909	Oversensing	30/04/2021	30/04/2021
									A0712	Pacing Problem	30/04/2021	30/04/2021
							E060101	Asystole	A070909	Oversensing	30/04/2021	30/04/2021
									A0712	Pacing Problem	30/04/2021	30/04/2021
							E0602	Cardiac Arrest	A070909	Oversensing	30/04/2021	30/04/2021
									A0712	Pacing Problem	30/04/2021	30/04/2021
							F08	Hospitalization or Prolonged Hospitalization	A070909	Oversensing	30/04/2021	30/04/2021
									A0712	Pacing Problem	30/04/2021	30/04/2021
							F1905	Device Revision or Replacement	A070909	Oversensing	30/04/2021	30/04/2021
									A0712	Pacing Problem	30/04/2021	30/04/2021
					CT2985	Active	E0112	Dizziness	A070909	Oversensing	30/04/2021	30/04/2021
									A0712	Pacing Problem	30/04/2021	30/04/2021
							E060101	Asystole	A070909	Oversensing	30/04/2021	30/04/2021
									A0712	Pacing Problem	30/04/2021	30/04/2021
							E0602	Cardiac Arrest	A070909	Oversensing	30/04/2021	30/04/2021
									A0712	Pacing Problem	30/04/2021	30/04/2021
							F08	Hospitalization or Prolonged Hospitalization	A070909	Oversensing	30/04/2021	30/04/2021
									A0712	Pacing Problem	30/04/2021	30/04/2021
							F1905	Device Revision or Replacement	A070909	Oversensing	30/04/2021	30/04/2021
									A0712	Pacing Problem	30/04/2021	30/04/2021
					CT3001	Implantable metal	E0112	Dizziness	A070909	Oversensing	30/04/2021	30/04/2021
									A0712	Pacing Problem	30/04/2021	30/04/2021
							E060101	Asystole	A070909	Oversensing	30/04/2021	30/04/2021
									A0712	Pacing Problem	30/04/2021	30/04/2021
							E0602	Cardiac Arrest	A070909	Oversensing	30/04/2021	30/04/2021
									A0712	Pacing Problem	30/04/2021	30/04/2021
							F08	Hospitalization or Prolonged Hospitalization	A070909	Oversensing	30/04/2021	30/04/2021
									A0712	Pacing Problem	30/04/2021	30/04/2021
							F1905	Device Revision or Replacement	A070909	Oversensing	30/04/2021	30/04/2021
									A0712	Pacing Problem	30/04/2021	30/04/2021
					CT3002	Implantable synthetic polymer	E0112	Dizziness	A070909	Oversensing	30/04/2021	30/04/2021
									A0712	Pacing Problem	30/04/2021	30/04/2021
							E060101	Asystole	A070909	Oversensing	30/04/2021	30/04/2021
									A0712	Pacing Problem	30/04/2021	30/04/2021
							E0602	Cardiac Arrest	A070909	Oversensing	30/04/2021	30/04/2021
									A0712	Pacing Problem	30/04/2021	30/04/2021
							F08	Hospitalization or Prolonged Hospitalization	A070909	Oversensing	30/04/2021	30/04/2021

									A0712	Pacing Problem	30/04/2021	30/04/2021
							F1905	Device Revision or Replacement	A070909	Oversensing	30/04/2021	30/04/2021
									A0712	Pacing Problem	30/04/2021	30/04/2021
					CT3006	Implantable inorganic material	E0112	Dizziness	A070909	Oversensing	30/04/2021	30/04/2021
									A0712	Pacing Problem	30/04/2021	30/04/2021
							E060101	Asystole	A070909	Oversensing	30/04/2021	30/04/2021
									A0712	Pacing Problem	30/04/2021	30/04/2021
							E0602	Cardiac Arrest	A070909	Oversensing	30/04/2021	30/04/2021
									A0712	Pacing Problem	30/04/2021	30/04/2021
							F08	Hospitalization or Prolonged Hospitalization	A070909	Oversensing	30/04/2021	30/04/2021
									A0712	Pacing Problem	30/04/2021	30/04/2021
							F1905	Device Revision or Replacement	A070909	Oversensing	30/04/2021	30/04/2021
									A0712	Pacing Problem	30/04/2021	30/04/2021
					CT314	Electrically powered	E0112	Dizziness	A070909	Oversensing	30/04/2021	30/04/2021
									A0712	Pacing Problem	30/04/2021	30/04/2021
							E060101	Asystole	A070909	Oversensing	30/04/2021	30/04/2021
									A0712	Pacing Problem	30/04/2021	30/04/2021
							E0602	Cardiac Arrest	A070909	Oversensing	30/04/2021	30/04/2021
									A0712	Pacing Problem	30/04/2021	30/04/2021
							F08	Hospitalization or Prolonged Hospitalization	A070909	Oversensing	30/04/2021	30/04/2021
									A0712	Pacing Problem	30/04/2021	30/04/2021
							F1905	Device Revision or Replacement	A070909	Oversensing	30/04/2021	30/04/2021
									A0712	Pacing Problem	30/04/2021	30/04/2021
					CT321	Long-term surgical invasive	E0112	Dizziness	A070909	Oversensing	30/04/2021	30/04/2021
									A0712	Pacing Problem	30/04/2021	30/04/2021
							E060101	Asystole	A070909	Oversensing	30/04/2021	30/04/2021
									A0712	Pacing Problem	30/04/2021	30/04/2021
							E0602	Cardiac Arrest	A070909	Oversensing	30/04/2021	30/04/2021
									A0712	Pacing Problem	30/04/2021	30/04/2021
							F08	Hospitalization or Prolonged Hospitalization	A070909	Oversensing	30/04/2021	30/04/2021
									A0712	Pacing Problem	30/04/2021	30/04/2021
							F1905	Device Revision or Replacement	A070909	Oversensing	30/04/2021	30/04/2021
									A0712	Pacing Problem	30/04/2021	30/04/2021
					CT449	Generators and associated devices	E0112	Dizziness	A070909	Oversensing	30/04/2021	30/04/2021
									A0712	Pacing Problem	30/04/2021	30/04/2021
							E060101	Asystole	A070909	Oversensing	30/04/2021	30/04/2021
									A0712	Pacing Problem	30/04/2021	30/04/2021
							E0602	Cardiac Arrest	A070909	Oversensing	30/04/2021	30/04/2021
									A0712	Pacing Problem	30/04/2021	30/04/2021
							F08	Hospitalization or Prolonged Hospitalization	A070909	Oversensing	30/04/2021	30/04/2021
									A0712	Pacing Problem	30/04/2021	30/04/2021
							F1905	Device Revision or Replacement	A070909	Oversensing	30/04/2021	30/04/2021
									A0712	Pacing Problem	30/04/2021	30/04/2021
					CT981	Single use	E0112	Dizziness	A070909	Oversensing	30/04/2021	30/04/2021
									A0712	Pacing Problem	30/04/2021	30/04/2021
							E060101	Asystole	A070909	Oversensing	30/04/2021	30/04/2021
									A0712	Pacing Problem	30/04/2021	30/04/2021
							E0602	Cardiac Arrest	A070909	Oversensing	30/04/2021	30/04/2021
									A0712	Pacing Problem	30/04/2021	30/04/2021
							F08	Hospitalization or Prolonged Hospitalization	A070909	Oversensing	30/04/2021	30/04/2021
									A0712	Pacing Problem	30/04/2021	30/04/2021
							F1905	Device Revision or Replacement	A070909	Oversensing	30/04/2021	30/04/2021
									A0712	Pacing Problem	30/04/2021	30/04/2021
					CT983	Surgical invasive	E0112	Dizziness	A070909	Oversensing	30/04/2021	30/04/2021
									A0712	Pacing Problem	30/04/2021	30/04/2021
							E060101	Asystole	A070909	Oversensing	30/04/2021	30/04/2021
									A0712	Pacing Problem	30/04/2021	30/04/2021
							E0602	Cardiac Arrest	A070909	Oversensing	30/04/2021	30/04/2021
									A0712	Pacing Problem	30/04/2021	30/04/2021
							F08	Hospitalization or Prolonged Hospitalization	A070909	Oversensing	30/04/2021	30/04/2021
									A0712	Pacing Problem	30/04/2021	30/04/2021
							F1905	Device Revision or Replacement	A070909	Oversensing	30/04/2021	30/04/2021
									A0712	Pacing Problem	30/04/2021	30/04/2021
11074	2021	All other reportable incidents	36241	Pacemaker/defibrillator lead	CT1684	Cardiac pacemakers and associated devices	E2104	Electric Shock	A070909	Oversensing		
									A071302	Inappropriate/Inadequate Shock/Stimulation		
							E2403	No Clinical Signs, Symptoms or Conditions	A070909	Oversensing		
									A071302	Inappropriate/Inadequate Shock/Stimulation		
							F08	Hospitalization or Prolonged Hospitalization	A070909	Oversensing		
									A071302	Inappropriate/Inadequate Shock/Stimulation		
							F23	Unexpected Medical Intervention	A070909	Oversensing		
									A071302	Inappropriate/Inadequate Shock/Stimulation		
11075	2021	All other reportable incidents	17882	Physiologic-monitoring defibrillation system	CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A1603	Failure of Device to Self-Test		
							F27	Problem Identified During Non-Clinical Procedure	A1603	Failure of Device to Self-Test		
11076	2021	All other reportable incidents	37265	Dual-chamber implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	12/05/2021	12/05/2021
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	12/05/2021	12/05/2021
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	12/05/2021	12/05/2021
11077	2021	All other reportable incidents	35852	Single-chamber implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
11078	2021	All other reportable incidents	47265	Dual-chamber implantable pacemaker, rate-responsive	CT141	Electrodes/Leads and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	11/05/2021	11/05/2021
							F1901	Additional Surgery	A070504	Premature Discharge of Battery	11/05/2021	11/05/2021
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	11/05/2021	11/05/2021
					CT145	Cardiovascular devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	11/05/2021	11/05/2021
							F1901	Additional Surgery	A070504	Premature Discharge of Battery	11/05/2021	11/05/2021
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	11/05/2021	11/05/2021
					CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	11/05/2021	11/05/2021
							F1901	Additional Surgery	A070504	Premature Discharge of Battery	11/05/2021	11/05/2021
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	11/05/2021	11/05/2021
					CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	11/05/2021	11/05/2021

							F1901	Additional Surgery	A070504	Premature Discharge of Battery	11/05/2021	11/05/2021
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	11/05/2021	11/05/2021
					CT233	Surgical	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	11/05/2021	11/05/2021
							F1901	Additional Surgery	A070504	Premature Discharge of Battery	11/05/2021	11/05/2021
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	11/05/2021	11/05/2021
					CT2406	Implantable	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	11/05/2021	11/05/2021
							F1901	Additional Surgery	A070504	Premature Discharge of Battery	11/05/2021	11/05/2021
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	11/05/2021	11/05/2021
					CT2407	Active implantable	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	11/05/2021	11/05/2021
							F1901	Additional Surgery	A070504	Premature Discharge of Battery	11/05/2021	11/05/2021
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	11/05/2021	11/05/2021
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	11/05/2021	11/05/2021
							F1901	Additional Surgery	A070504	Premature Discharge of Battery	11/05/2021	11/05/2021
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	11/05/2021	11/05/2021
					CT2727	Active implant devices/ancillaries	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	11/05/2021	11/05/2021
							F1901	Additional Surgery	A070504	Premature Discharge of Battery	11/05/2021	11/05/2021
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	11/05/2021	11/05/2021
					CT2750	Implantable cardiac electrical stimulators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	11/05/2021	11/05/2021
							F1901	Additional Surgery	A070504	Premature Discharge of Battery	11/05/2021	11/05/2021
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	11/05/2021	11/05/2021
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	11/05/2021	11/05/2021
							F1901	Additional Surgery	A070504	Premature Discharge of Battery	11/05/2021	11/05/2021
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	11/05/2021	11/05/2021
					CT2985	Active	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	11/05/2021	11/05/2021
							F1901	Additional Surgery	A070504	Premature Discharge of Battery	11/05/2021	11/05/2021
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	11/05/2021	11/05/2021
					CT3001	Implantable metal	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	11/05/2021	11/05/2021
							F1901	Additional Surgery	A070504	Premature Discharge of Battery	11/05/2021	11/05/2021
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	11/05/2021	11/05/2021
					CT3002	Implantable synthetic polymer	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	11/05/2021	11/05/2021
							F1901	Additional Surgery	A070504	Premature Discharge of Battery	11/05/2021	11/05/2021
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	11/05/2021	11/05/2021
					CT3006	Implantable inorganic material	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	11/05/2021	11/05/2021
							F1901	Additional Surgery	A070504	Premature Discharge of Battery	11/05/2021	11/05/2021
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	11/05/2021	11/05/2021
					CT314	Electrically powered	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	11/05/2021	11/05/2021
							F1901	Additional Surgery	A070504	Premature Discharge of Battery	11/05/2021	11/05/2021
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	11/05/2021	11/05/2021
					CT321	Long-term surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	11/05/2021	11/05/2021
							F1901	Additional Surgery	A070504	Premature Discharge of Battery	11/05/2021	11/05/2021
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	11/05/2021	11/05/2021
					CT449	Generators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	11/05/2021	11/05/2021
							F1901	Additional Surgery	A070504	Premature Discharge of Battery	11/05/2021	11/05/2021
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	11/05/2021	11/05/2021
					CT981	Single use	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	11/05/2021	11/05/2021
							F1901	Additional Surgery	A070504	Premature Discharge of Battery	11/05/2021	11/05/2021
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	11/05/2021	11/05/2021
					CT983	Surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	11/05/2021	11/05/2021
							F1901	Additional Surgery	A070504	Premature Discharge of Battery	11/05/2021	11/05/2021
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	11/05/2021	11/05/2021
11079	2021	All other reportable incidents	47265	Dual-chamber implantable pacemaker, rate-responsive	CT141	Electrodes/Leads and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT145	Cardiovascular devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT233	Surgical	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT2406	Implantable	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT2407	Active implantable	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT2727	Active implant devices/ancillaries	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT2750	Implantable cardiac electrical stimulators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT2985	Active	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT3001	Implantable metal	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT3002	Implantable synthetic polymer	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT3006	Implantable inorganic material	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT314	Electrically powered	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT321	Long-term surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT449	Generators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT981	Single use	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		

					CT983	Surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
11080	2021	All other reportable incidents	35852	Single-chamber implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E2104	Electric Shock	A070909	Oversensing		
									A071302	Inappropriate/Inadequate Shock/Stimulation		
							F23	Unexpected Medical Intervention	A070909	Oversensing		
									A071302	Inappropriate/Inadequate Shock/Stimulation		
11081	2021	All other reportable incidents	36241	Pacemaker/defibrillator lead	CT1684	Cardiac pacemakers and associated devices	E2104	Electric Shock	A070909	Oversensing		
									A071302	Inappropriate/Inadequate Shock/Stimulation		
							F1905	Device Revision or Replacement	A070909	Oversensing		
									A071302	Inappropriate/Inadequate Shock/Stimulation		
							F23	Unexpected Medical Intervention	A070909	Oversensing		
									A071302	Inappropriate/Inadequate Shock/Stimulation		
11082	2021	All other reportable incidents	17882	Physiologic-monitoring defibrillation system	CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070803	Failure to Power Up		
							F27	Problem Identified During Non-Clinical Procedure	A070803	Failure to Power Up		
11083	2021	All other reportable incidents	17882	Physiologic-monitoring defibrillation system	CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A0719	Unexpected Shutdown		
							F26	No Health Consequences or Impact	A0719	Unexpected Shutdown		
11084	2021	All other reportable incidents	35223	Endocardial/interventricular septal pacing lead	CT1684	Cardiac pacemakers and associated devices	E0605	Cardiac Tamponade	A27	Appropriate Device Problem Term/Code not Available	07/05/2021	07/05/2021
							E0619	Pericardial Effusion	A27	Appropriate Device Problem Term/Code not Available	07/05/2021	07/05/2021
							F08	Hospitalization or Prolonged Hospitalization	A27	Appropriate Device Problem Term/Code not Available	07/05/2021	07/05/2021
							F12	Serious Injury/ Illness/ Impairment	A27	Appropriate Device Problem Term/Code not Available	07/05/2021	07/05/2021
							F1905	Device Revision or Replacement	A27	Appropriate Device Problem Term/Code not Available	07/05/2021	07/05/2021
11085	2021	All other reportable incidents	36241	Pacemaker/defibrillator lead	CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A072201	High Impedance		
							F26	No Health Consequences or Impact	A072201	High Impedance		
11086	2021	All other reportable incidents	47265	Dual-chamber implantable pacemaker, rate-responsive	CT141	Electrodes/Leads and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070909	Oversensing		
									A072201	High Impedance		
									A090801	Signal Artifact/Noise		
							F15	Recognized Device or Procedural Complication	A070909	Oversensing		
									A072201	High Impedance		
									A090801	Signal Artifact/Noise		
					CT145	Cardiovascular devices	E2403	No Clinical Signs, Symptoms or Conditions	A070909	Oversensing		
									A072201	High Impedance		
									A090801	Signal Artifact/Noise		
							F15	Recognized Device or Procedural Complication	A070909	Oversensing		
									A072201	High Impedance		
									A090801	Signal Artifact/Noise		
					CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070909	Oversensing		
									A072201	High Impedance		
									A090801	Signal Artifact/Noise		
							F15	Recognized Device or Procedural Complication	A070909	Oversensing		
									A072201	High Impedance		
									A090801	Signal Artifact/Noise		
					CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A070909	Oversensing		
									A072201	High Impedance		
									A090801	Signal Artifact/Noise		
							F15	Recognized Device or Procedural Complication	A070909	Oversensing		
									A072201	High Impedance		
									A090801	Signal Artifact/Noise		
					CT233	Surgical	E2403	No Clinical Signs, Symptoms or Conditions	A070909	Oversensing		
									A072201	High Impedance		
									A090801	Signal Artifact/Noise		
							F15	Recognized Device or Procedural Complication	A070909	Oversensing		
									A072201	High Impedance		
									A090801	Signal Artifact/Noise		
					CT2406	Implantable	E2403	No Clinical Signs, Symptoms or Conditions	A070909	Oversensing		
									A072201	High Impedance		
									A090801	Signal Artifact/Noise		
							F15	Recognized Device or Procedural Complication	A070909	Oversensing		
									A072201	High Impedance		
									A090801	Signal Artifact/Noise		
					CT2407	Active implantable	E2403	No Clinical Signs, Symptoms or Conditions	A070909	Oversensing		
									A072201	High Impedance		
									A090801	Signal Artifact/Noise		
							F15	Recognized Device or Procedural Complication	A070909	Oversensing		
									A072201	High Impedance		
									A090801	Signal Artifact/Noise		
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A070909	Oversensing		
									A072201	High Impedance		
									A090801	Signal Artifact/Noise		
							F15	Recognized Device or Procedural Complication	A070909	Oversensing		
									A072201	High Impedance		
									A090801	Signal Artifact/Noise		
					CT2727	Active implant devices/ancillaries	E2403	No Clinical Signs, Symptoms or Conditions	A070909	Oversensing		
									A072201	High Impedance		
									A090801	Signal Artifact/Noise		
							F15	Recognized Device or Procedural Complication	A070909	Oversensing		
									A072201	High Impedance		
									A090801	Signal Artifact/Noise		
					CT2750	Implantable cardiac electrical stimulators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070909	Oversensing		
									A072201	High Impedance		
									A090801	Signal Artifact/Noise		
							F15	Recognized Device or Procedural Complication	A070909	Oversensing		
									A072201	High Impedance		
									A090801	Signal Artifact/Noise		
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A070909	Oversensing		

									A072201	High Impedance		
									A090801	Signal Artifact/Noise		
							F15	Recognized Device or Procedural Complication	A070909	Oversensing		
									A072201	High Impedance		
									A090801	Signal Artifact/Noise		
					CT2985	Active	E2403	No Clinical Signs, Symptoms or Conditions	A070909	Oversensing		
									A072201	High Impedance		
									A090801	Signal Artifact/Noise		
							F15	Recognized Device or Procedural Complication	A070909	Oversensing		
									A072201	High Impedance		
									A090801	Signal Artifact/Noise		
					CT3001	Implantable metal	E2403	No Clinical Signs, Symptoms or Conditions	A070909	Oversensing		
									A072201	High Impedance		
									A090801	Signal Artifact/Noise		
							F15	Recognized Device or Procedural Complication	A070909	Oversensing		
									A072201	High Impedance		
									A090801	Signal Artifact/Noise		
					CT3002	Implantable synthetic polymer	E2403	No Clinical Signs, Symptoms or Conditions	A070909	Oversensing		
									A072201	High Impedance		
									A090801	Signal Artifact/Noise		
							F15	Recognized Device or Procedural Complication	A070909	Oversensing		
									A072201	High Impedance		
									A090801	Signal Artifact/Noise		
					CT3006	Implantable inorganic material	E2403	No Clinical Signs, Symptoms or Conditions	A070909	Oversensing		
									A072201	High Impedance		
									A090801	Signal Artifact/Noise		
							F15	Recognized Device or Procedural Complication	A070909	Oversensing		
									A072201	High Impedance		
									A090801	Signal Artifact/Noise		
					CT314	Electrically powered	E2403	No Clinical Signs, Symptoms or Conditions	A070909	Oversensing		
									A072201	High Impedance		
									A090801	Signal Artifact/Noise		
							F15	Recognized Device or Procedural Complication	A070909	Oversensing		
									A072201	High Impedance		
									A090801	Signal Artifact/Noise		
					CT321	Long-term surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A070909	Oversensing		
									A072201	High Impedance		
									A090801	Signal Artifact/Noise		
							F15	Recognized Device or Procedural Complication	A070909	Oversensing		
									A072201	High Impedance		
									A090801	Signal Artifact/Noise		
					CT449	Generators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070909	Oversensing		
									A072201	High Impedance		
									A090801	Signal Artifact/Noise		
							F15	Recognized Device or Procedural Complication	A070909	Oversensing		
									A072201	High Impedance		
									A090801	Signal Artifact/Noise		
					CT981	Single use	E2403	No Clinical Signs, Symptoms or Conditions	A070909	Oversensing		
									A072201	High Impedance		
									A090801	Signal Artifact/Noise		
							F15	Recognized Device or Procedural Complication	A070909	Oversensing		
									A072201	High Impedance		
									A090801	Signal Artifact/Noise		
					CT983	Surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A070909	Oversensing		
									A072201	High Impedance		
									A090801	Signal Artifact/Noise		
							F15	Recognized Device or Procedural Complication	A070909	Oversensing		
									A072201	High Impedance		
									A090801	Signal Artifact/Noise		
11087	2021	All other reportable incidents	17882	Physiologic-monitoring defibrillation system	CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070908	Failure to Sense		
							F27	Problem Identified During Non-Clinical Procedure	A070908	Failure to Sense		
11088	2021	All other reportable incidents	35852	Single-chamber implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E2104	Electric Shock	A070909	Oversensing		
									A071302	Inappropriate/Inadequate Shock/Stimulation		
									A090801	Signal Artifact/Noise		
							E2403	No Clinical Signs, Symptoms or Conditions	A070909	Oversensing		
									A071302	Inappropriate/Inadequate Shock/Stimulation		
									A090801	Signal Artifact/Noise		
							F26	No Health Consequences or Impact	A070909	Oversensing		
									A071302	Inappropriate/Inadequate Shock/Stimulation		
									A090801	Signal Artifact/Noise		
11089	2021	All other reportable incidents	35852	Single-chamber implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A070909	Oversensing	18/05/2021	18/05/2021
									A090801	Signal Artifact/Noise	18/05/2021	18/05/2021
									A1301	Failure to Read Input Signal	18/05/2021	18/05/2021
							F08	Hospitalization or Prolonged Hospitalization	A070909	Oversensing	18/05/2021	18/05/2021
									A090801	Signal Artifact/Noise	18/05/2021	18/05/2021
									A1301	Failure to Read Input Signal	18/05/2021	18/05/2021
							F12	Serious Injury/ Illness/ Impairment	A070909	Oversensing	18/05/2021	18/05/2021
									A090801	Signal Artifact/Noise	18/05/2021	18/05/2021
									A1301	Failure to Read Input Signal	18/05/2021	18/05/2021
							F19	Surgical Intervention	A070909	Oversensing	18/05/2021	18/05/2021
									A090801	Signal Artifact/Noise	18/05/2021	18/05/2021
									A1301	Failure to Read Input Signal	18/05/2021	18/05/2021
							F1905	Device Revision or Replacement	A070909	Oversensing	18/05/2021	18/05/2021
									A090801	Signal Artifact/Noise	18/05/2021	18/05/2021
									A1301	Failure to Read Input Signal	18/05/2021	18/05/2021
11090	2021	All other reportable incidents	36241	Pacemaker/defibrillator lead	CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070909	Oversensing	18/05/2021	18/05/2021
									A090801	Signal Artifact/Noise	18/05/2021	18/05/2021
									A1301	Failure to Read Input Signal	18/05/2021	18/05/2021

							F08	Hospitalization or Prolonged Hospitalization	A070909	Oversensing	18/05/2021	18/05/2021
									A090801	Signal Artifact/Noise	18/05/2021	18/05/2021
									A1301	Failure to Read Input Signal	18/05/2021	18/05/2021
							F12	Serious Injury/ Illness/ Impairment	A070909	Oversensing	18/05/2021	18/05/2021
									A090801	Signal Artifact/Noise	18/05/2021	18/05/2021
									A1301	Failure to Read Input Signal	18/05/2021	18/05/2021
							F19	Surgical Intervention	A070909	Oversensing	18/05/2021	18/05/2021
									A090801	Signal Artifact/Noise	18/05/2021	18/05/2021
									A1301	Failure to Read Input Signal	18/05/2021	18/05/2021
11091	2021	All other reportable incidents	35852	Single-chamber implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A070909	Oversensing		
							F26	No Health Consequences or Impact	A070909	Oversensing		
11092	2021	All other reportable incidents	47270	Cardiac resynchronization therapy implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	11/05/2021	11/05/2021
							F1901	Additional Surgery	A070504	Premature Discharge of Battery	11/05/2021	11/05/2021
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	11/05/2021	11/05/2021
11093	2021	All other reportable incidents	35852	Single-chamber implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A070603	Failure to Charge		
									A160503	Inappropriate or Unexpected Reset		
							F23	Unexpected Medical Intervention	A070603	Failure to Charge		
									A160503	Inappropriate or Unexpected Reset		
11094	2021	All other reportable incidents	47263	Cardiac resynchronization therapy implantable pacemaker	CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	16/06/2021	16/06/2021
							F1901	Additional Surgery	A070504	Premature Discharge of Battery	16/06/2021	16/06/2021
					CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	16/06/2021	16/06/2021
							F1901	Additional Surgery	A070504	Premature Discharge of Battery	16/06/2021	16/06/2021
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	16/06/2021	16/06/2021
							F1901	Additional Surgery	A070504	Premature Discharge of Battery	16/06/2021	16/06/2021
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	16/06/2021	16/06/2021
							F1901	Additional Surgery	A070504	Premature Discharge of Battery	16/06/2021	16/06/2021
11095	2021	All other reportable incidents	47270	Cardiac resynchronization therapy implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	24/05/2021	24/05/2021
							F1901	Additional Surgery	A070504	Premature Discharge of Battery	24/05/2021	24/05/2021
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	24/05/2021	24/05/2021
11096	2021	All other reportable incidents	47263	Cardiac resynchronization therapy implantable pacemaker	CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070101	Failure to Capture		
							F26	No Health Consequences or Impact	A070101	Failure to Capture		
					CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A070101	Failure to Capture		
							F26	No Health Consequences or Impact	A070101	Failure to Capture		
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A070101	Failure to Capture		
							F26	No Health Consequences or Impact	A070101	Failure to Capture		
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A070101	Failure to Capture		
							F26	No Health Consequences or Impact	A070101	Failure to Capture		
11097	2021	All other reportable incidents	35223	Endocardial/interventricular septal pacing lead	CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A05	Mechanical Problem		
									A070909	Oversensing		
									A072201	High Impedance		
									A090801	Signal Artifact/Noise		
							F26	No Health Consequences or Impact	A05	Mechanical Problem		
									A070909	Oversensing		
									A072201	High Impedance		
									A090801	Signal Artifact/Noise		
11098	2021	All other reportable incidents	35223	Endocardial/interventricular septal pacing lead	CT1684	Cardiac pacemakers and associated devices	E0112	Dizziness	A070104	Unstable Capture Threshold		
									A070909	Oversensing		
									A0712	Pacing Problem		
							E060104	Bradycardia	A070104	Unstable Capture Threshold		
									A070909	Oversensing		
									A0712	Pacing Problem		
							E2403	No Clinical Signs, Symptoms or Conditions	A070104	Unstable Capture Threshold		
									A070909	Oversensing		
									A0712	Pacing Problem		
							F23	Unexpected Medical Intervention	A070104	Unstable Capture Threshold		
									A070909	Oversensing		
									A0712	Pacing Problem		
11099	2021	All other reportable incidents	35852	Single-chamber implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E2401	Insufficient Information	A0705	Battery Problem	12/05/2021	12/05/2021
									A070504	Premature Discharge of Battery	12/05/2021	12/05/2021
							E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem	12/05/2021	12/05/2021
									A070504	Premature Discharge of Battery	12/05/2021	12/05/2021
							F08	Hospitalization or Prolonged Hospitalization	A0705	Battery Problem	12/05/2021	12/05/2021
									A070504	Premature Discharge of Battery	12/05/2021	12/05/2021
							F19	Surgical Intervention	A0705	Battery Problem	12/05/2021	12/05/2021
									A070504	Premature Discharge of Battery	12/05/2021	12/05/2021
11100	2021	All other reportable incidents	35223	Endocardial/interventricular septal pacing lead	CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A0401	Break	24/08/2021	24/08/2021
									A07	Electrical /Electronic Property Problem	24/08/2021	24/08/2021
									A070909	Oversensing	24/08/2021	24/08/2021
									A072202	Low Impedance	24/08/2021	24/08/2021
									A090801	Signal Artifact/Noise	24/08/2021	24/08/2021
							F1905	Device Revision or Replacement	A0401	Break	24/08/2021	24/08/2021
									A07	Electrical /Electronic Property Problem	24/08/2021	24/08/2021
									A070909	Oversensing	24/08/2021	24/08/2021
									A072202	Low Impedance	24/08/2021	24/08/2021
									A090801	Signal Artifact/Noise	24/08/2021	24/08/2021
							F26	No Health Consequences or Impact	A0401	Break	24/08/2021	24/08/2021
									A07	Electrical /Electronic Property Problem	24/08/2021	24/08/2021
									A070909	Oversensing	24/08/2021	24/08/2021
									A072202	Low Impedance	24/08/2021	24/08/2021
									A090801	Signal Artifact/Noise	24/08/2021	24/08/2021
11101	2021	All other reportable incidents	37265	Dual-chamber implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	19/05/2021	19/05/2021
							F1901	Additional Surgery	A070504	Premature Discharge of Battery	19/05/2021	19/05/2021
11102	2021	All other reportable incidents	47265	Dual-chamber implantable pacemaker, rate-responsive	CT141	Electrodes/Leads and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	14/04/2022	14/04/2022
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	14/04/2022	14/04/2022
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	14/04/2022	14/04/2022
					CT145	Cardiovascular devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	14/04/2022	14/04/2022
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	14/04/2022	14/04/2022
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	14/04/2022	14/04/2022

					CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	14/04/2022	14/04/2022
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	14/04/2022	14/04/2022
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	14/04/2022	14/04/2022
					CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	14/04/2022	14/04/2022
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	14/04/2022	14/04/2022
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	14/04/2022	14/04/2022
					CT233	Surgical	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	14/04/2022	14/04/2022
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	14/04/2022	14/04/2022
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	14/04/2022	14/04/2022
					CT2406	Implantable	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	14/04/2022	14/04/2022
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	14/04/2022	14/04/2022
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	14/04/2022	14/04/2022
					CT2407	Active implantable	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	14/04/2022	14/04/2022
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	14/04/2022	14/04/2022
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	14/04/2022	14/04/2022
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	14/04/2022	14/04/2022
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	14/04/2022	14/04/2022
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	14/04/2022	14/04/2022
					CT2727	Active implant devices/ancillaries	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	14/04/2022	14/04/2022
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	14/04/2022	14/04/2022
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	14/04/2022	14/04/2022
					CT2750	Implantable cardiac electrical stimulators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	14/04/2022	14/04/2022
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	14/04/2022	14/04/2022
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	14/04/2022	14/04/2022
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	14/04/2022	14/04/2022
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	14/04/2022	14/04/2022
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	14/04/2022	14/04/2022
					CT2985	Active	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	14/04/2022	14/04/2022
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	14/04/2022	14/04/2022
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	14/04/2022	14/04/2022
					CT3001	Implantable metal	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	14/04/2022	14/04/2022
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	14/04/2022	14/04/2022
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	14/04/2022	14/04/2022
					CT3002	Implantable synthetic polymer	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	14/04/2022	14/04/2022
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	14/04/2022	14/04/2022
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	14/04/2022	14/04/2022
					CT3006	Implantable inorganic material	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	14/04/2022	14/04/2022
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	14/04/2022	14/04/2022
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	14/04/2022	14/04/2022
					CT314	Electrically powered	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	14/04/2022	14/04/2022
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	14/04/2022	14/04/2022
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	14/04/2022	14/04/2022
					CT321	Long-term surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	14/04/2022	14/04/2022
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	14/04/2022	14/04/2022
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	14/04/2022	14/04/2022
					CT449	Generators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	14/04/2022	14/04/2022
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	14/04/2022	14/04/2022
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	14/04/2022	14/04/2022
					CT981	Single use	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	14/04/2022	14/04/2022
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	14/04/2022	14/04/2022
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	14/04/2022	14/04/2022
					CT983	Surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	14/04/2022	14/04/2022
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	14/04/2022	14/04/2022
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	14/04/2022	14/04/2022
11103	2021	All other reportable incidents	35223	Endocardial/interventricular septal pacing lead	CT1684	Cardiac pacemakers and associated devices	E0717	Dyspnea	A0401	Break		
									A07	Electrical /Electronic Property Problem		
									A070909	Oversensing		
									A0712	Pacing Problem		
									A0722	Impedance Problem		
									A072201	High Impedance		
							E2312	Fatigue	A0401	Break		
									A07	Electrical /Electronic Property Problem		
									A070909	Oversensing		
									A0712	Pacing Problem		
									A0722	Impedance Problem		
									A072201	High Impedance		
							E2403	No Clinical Signs, Symptoms or Conditions	A0401	Break		
									A07	Electrical /Electronic Property Problem		
									A070909	Oversensing		
									A0712	Pacing Problem		
									A0722	Impedance Problem		
									A072201	High Impedance		
							F23	Unexpected Medical Intervention	A0401	Break		
									A07	Electrical /Electronic Property Problem		
									A070909	Oversensing		
									A0712	Pacing Problem		
									A0722	Impedance Problem		
									A072201	High Impedance		
							F26	No Health Consequences or Impact	A0401	Break		
									A07	Electrical /Electronic Property Problem		
									A070909	Oversensing		
									A0712	Pacing Problem		
									A0722	Impedance Problem		
									A072201	High Impedance		
11104	2021	All other reportable incidents	47265	Dual-chamber implantable pacemaker, rate-responsive	CT141	Electrodes/Leads and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem	07/06/2021	07/06/2021
									A070504	Premature Discharge of Battery	07/06/2021	07/06/2021
							F08	Hospitalization or Prolonged Hospitalization	A0705	Battery Problem	07/06/2021	07/06/2021
									A070504	Premature Discharge of Battery	07/06/2021	07/06/2021

		F1905	Device Revision or Replacement	A0705	Battery Problem	07/06/2021	07/06/2021
				A070504	Premature Discharge of Battery	07/06/2021	07/06/2021
CT145	Cardiovascular devices	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem	07/06/2021	07/06/2021
				A070504	Premature Discharge of Battery	07/06/2021	07/06/2021
		F08	Hospitalization or Prolonged Hospitalization	A0705	Battery Problem	07/06/2021	07/06/2021
				A070504	Premature Discharge of Battery	07/06/2021	07/06/2021
		F1905	Device Revision or Replacement	A0705	Battery Problem	07/06/2021	07/06/2021
				A070504	Premature Discharge of Battery	07/06/2021	07/06/2021
CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem	07/06/2021	07/06/2021
				A070504	Premature Discharge of Battery	07/06/2021	07/06/2021
		F08	Hospitalization or Prolonged Hospitalization	A0705	Battery Problem	07/06/2021	07/06/2021
				A070504	Premature Discharge of Battery	07/06/2021	07/06/2021
		F1905	Device Revision or Replacement	A0705	Battery Problem	07/06/2021	07/06/2021
				A070504	Premature Discharge of Battery	07/06/2021	07/06/2021
CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem	07/06/2021	07/06/2021
				A070504	Premature Discharge of Battery	07/06/2021	07/06/2021
		F08	Hospitalization or Prolonged Hospitalization	A0705	Battery Problem	07/06/2021	07/06/2021
				A070504	Premature Discharge of Battery	07/06/2021	07/06/2021
		F1905	Device Revision or Replacement	A0705	Battery Problem	07/06/2021	07/06/2021
				A070504	Premature Discharge of Battery	07/06/2021	07/06/2021
CT233	Surgical	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem	07/06/2021	07/06/2021
				A070504	Premature Discharge of Battery	07/06/2021	07/06/2021
		F08	Hospitalization or Prolonged Hospitalization	A0705	Battery Problem	07/06/2021	07/06/2021
				A070504	Premature Discharge of Battery	07/06/2021	07/06/2021
		F1905	Device Revision or Replacement	A0705	Battery Problem	07/06/2021	07/06/2021
				A070504	Premature Discharge of Battery	07/06/2021	07/06/2021
CT2406	Implantable	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem	07/06/2021	07/06/2021
				A070504	Premature Discharge of Battery	07/06/2021	07/06/2021
		F08	Hospitalization or Prolonged Hospitalization	A0705	Battery Problem	07/06/2021	07/06/2021
				A070504	Premature Discharge of Battery	07/06/2021	07/06/2021
		F1905	Device Revision or Replacement	A0705	Battery Problem	07/06/2021	07/06/2021
				A070504	Premature Discharge of Battery	07/06/2021	07/06/2021
CT2407	Active implantable	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem	07/06/2021	07/06/2021
				A070504	Premature Discharge of Battery	07/06/2021	07/06/2021
		F08	Hospitalization or Prolonged Hospitalization	A0705	Battery Problem	07/06/2021	07/06/2021
				A070504	Premature Discharge of Battery	07/06/2021	07/06/2021
		F1905	Device Revision or Replacement	A0705	Battery Problem	07/06/2021	07/06/2021
				A070504	Premature Discharge of Battery	07/06/2021	07/06/2021
CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem	07/06/2021	07/06/2021
				A070504	Premature Discharge of Battery	07/06/2021	07/06/2021
		F08	Hospitalization or Prolonged Hospitalization	A0705	Battery Problem	07/06/2021	07/06/2021
				A070504	Premature Discharge of Battery	07/06/2021	07/06/2021
		F1905	Device Revision or Replacement	A0705	Battery Problem	07/06/2021	07/06/2021
				A070504	Premature Discharge of Battery	07/06/2021	07/06/2021
CT2727	Active implant devices/ancillaries	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem	07/06/2021	07/06/2021
				A070504	Premature Discharge of Battery	07/06/2021	07/06/2021
		F08	Hospitalization or Prolonged Hospitalization	A0705	Battery Problem	07/06/2021	07/06/2021
				A070504	Premature Discharge of Battery	07/06/2021	07/06/2021
		F1905	Device Revision or Replacement	A0705	Battery Problem	07/06/2021	07/06/2021
				A070504	Premature Discharge of Battery	07/06/2021	07/06/2021
CT2750	Implantable cardiac electrical stimulators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem	07/06/2021	07/06/2021
				A070504	Premature Discharge of Battery	07/06/2021	07/06/2021
		F08	Hospitalization or Prolonged Hospitalization	A0705	Battery Problem	07/06/2021	07/06/2021
				A070504	Premature Discharge of Battery	07/06/2021	07/06/2021
		F1905	Device Revision or Replacement	A0705	Battery Problem	07/06/2021	07/06/2021
				A070504	Premature Discharge of Battery	07/06/2021	07/06/2021
CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem	07/06/2021	07/06/2021
				A070504	Premature Discharge of Battery	07/06/2021	07/06/2021
		F08	Hospitalization or Prolonged Hospitalization	A0705	Battery Problem	07/06/2021	07/06/2021
				A070504	Premature Discharge of Battery	07/06/2021	07/06/2021
		F1905	Device Revision or Replacement	A0705	Battery Problem	07/06/2021	07/06/2021
				A070504	Premature Discharge of Battery	07/06/2021	07/06/2021
CT2985	Active	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem	07/06/2021	07/06/2021
				A070504	Premature Discharge of Battery	07/06/2021	07/06/2021
		F08	Hospitalization or Prolonged Hospitalization	A0705	Battery Problem	07/06/2021	07/06/2021
				A070504	Premature Discharge of Battery	07/06/2021	07/06/2021
		F1905	Device Revision or Replacement	A0705	Battery Problem	07/06/2021	07/06/2021
				A070504	Premature Discharge of Battery	07/06/2021	07/06/2021
CT3001	Implantable metal	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem	07/06/2021	07/06/2021
				A070504	Premature Discharge of Battery	07/06/2021	07/06/2021
		F08	Hospitalization or Prolonged Hospitalization	A0705	Battery Problem	07/06/2021	07/06/2021
				A070504	Premature Discharge of Battery	07/06/2021	07/06/2021
		F1905	Device Revision or Replacement	A0705	Battery Problem	07/06/2021	07/06/2021
				A070504	Premature Discharge of Battery	07/06/2021	07/06/2021
CT3002	Implantable synthetic polymer	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem	07/06/2021	07/06/2021
				A070504	Premature Discharge of Battery	07/06/2021	07/06/2021
		F08	Hospitalization or Prolonged Hospitalization	A0705	Battery Problem	07/06/2021	07/06/2021
				A070504	Premature Discharge of Battery	07/06/2021	07/06/2021
		F1905	Device Revision or Replacement	A0705	Battery Problem	07/06/2021	07/06/2021
				A070504	Premature Discharge of Battery	07/06/2021	07/06/2021
CT3006	Implantable inorganic material	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem	07/06/2021	07/06/2021
				A070504	Premature Discharge of Battery	07/06/2021	07/06/2021
		F08	Hospitalization or Prolonged Hospitalization	A0705	Battery Problem	07/06/2021	07/06/2021
				A070504	Premature Discharge of Battery	07/06/2021	07/06/2021
		F1905	Device Revision or Replacement	A0705	Battery Problem	07/06/2021	07/06/2021
				A070504	Premature Discharge of Battery	07/06/2021	07/06/2021
CT314	Electrically powered	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem	07/06/2021	07/06/2021
				A070504	Premature Discharge of Battery	07/06/2021	07/06/2021

							F08	Hospitalization or Prolonged Hospitalization	A0705	Battery Problem	07/06/2021	07/06/2021
									A070504	Premature Discharge of Battery	07/06/2021	07/06/2021
							F1905	Device Revision or Replacement	A0705	Battery Problem	07/06/2021	07/06/2021
									A070504	Premature Discharge of Battery	07/06/2021	07/06/2021
					CT321	Long-term surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem	07/06/2021	07/06/2021
									A070504	Premature Discharge of Battery	07/06/2021	07/06/2021
							F08	Hospitalization or Prolonged Hospitalization	A0705	Battery Problem	07/06/2021	07/06/2021
									A070504	Premature Discharge of Battery	07/06/2021	07/06/2021
							F1905	Device Revision or Replacement	A0705	Battery Problem	07/06/2021	07/06/2021
									A070504	Premature Discharge of Battery	07/06/2021	07/06/2021
					CT449	Generators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem	07/06/2021	07/06/2021
									A070504	Premature Discharge of Battery	07/06/2021	07/06/2021
							F08	Hospitalization or Prolonged Hospitalization	A0705	Battery Problem	07/06/2021	07/06/2021
									A070504	Premature Discharge of Battery	07/06/2021	07/06/2021
							F1905	Device Revision or Replacement	A0705	Battery Problem	07/06/2021	07/06/2021
									A070504	Premature Discharge of Battery	07/06/2021	07/06/2021
					CT981	Single use	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem	07/06/2021	07/06/2021
									A070504	Premature Discharge of Battery	07/06/2021	07/06/2021
							F08	Hospitalization or Prolonged Hospitalization	A0705	Battery Problem	07/06/2021	07/06/2021
									A070504	Premature Discharge of Battery	07/06/2021	07/06/2021
							F1905	Device Revision or Replacement	A0705	Battery Problem	07/06/2021	07/06/2021
									A070504	Premature Discharge of Battery	07/06/2021	07/06/2021
					CT983	Surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem	07/06/2021	07/06/2021
									A070504	Premature Discharge of Battery	07/06/2021	07/06/2021
							F08	Hospitalization or Prolonged Hospitalization	A0705	Battery Problem	07/06/2021	07/06/2021
									A070504	Premature Discharge of Battery	07/06/2021	07/06/2021
							F1905	Device Revision or Replacement	A0705	Battery Problem	07/06/2021	07/06/2021
									A070504	Premature Discharge of Battery	07/06/2021	07/06/2021
11105	2021	All other reportable incidents	36241	Pacemaker/defibrillator lead	CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A072201	High Impedance		
							F08	Hospitalization or Prolonged Hospitalization	A072201	High Impedance		
							F1904	Device Repositioning	A072201	High Impedance		
11106	2021	All other reportable incidents	35852	Single-chamber implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E2104	Electric Shock	A071302	Inappropriate/Inadequate Shock/Stimulation		
							E2403	No Clinical Signs, Symptoms or Conditions	A071302	Inappropriate/Inadequate Shock/Stimulation		
							F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A071302	Inappropriate/Inadequate Shock/Stimulation		
11107	2021	All other reportable incidents	36241	Pacemaker/defibrillator lead	CT1684	Cardiac pacemakers and associated devices	E0112	Dizziness	A070909	Oversensing	12/05/2021	12/05/2021
									A0712	Pacing Problem	12/05/2021	12/05/2021
							F08	Hospitalization or Prolonged Hospitalization	A070909	Oversensing	12/05/2021	12/05/2021
									A0712	Pacing Problem	12/05/2021	12/05/2021
							F12	Serious Injury/ Illness/ Impairment	A070909	Oversensing	12/05/2021	12/05/2021
									A0712	Pacing Problem	12/05/2021	12/05/2021
							F1905	Device Revision or Replacement	A070909	Oversensing	12/05/2021	12/05/2021
									A0712	Pacing Problem	12/05/2021	12/05/2021
11108	2021	All other reportable incidents	36241	Pacemaker/defibrillator lead	CT1684	Cardiac pacemakers and associated devices	E0604	Cardiac Perforation	A051201	Device Dislodged or Dislocated		
									A070101	Failure to Capture		
							E2103	Device Overstimulation of Tissue	A051201	Device Dislodged or Dislocated		
									A070101	Failure to Capture		
							E2114	Perforation	A051201	Device Dislodged or Dislocated		
									A070101	Failure to Capture		
							E2330	Pain	A051201	Device Dislodged or Dislocated		
									A070101	Failure to Capture		
							F1905	Device Revision or Replacement	A051201	Device Dislodged or Dislocated		
									A070101	Failure to Capture		
							F2203	Imaging Required	A051201	Device Dislodged or Dislocated		
									A070101	Failure to Capture		
11109	2021	All other reportable incidents	47265	Dual-chamber implantable pacemaker, rate-responsive	CT141	Electrodes/Leads and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F11	Non-Serious Injury/ Illness/ Impairment	A070504	Premature Discharge of Battery		
					CT145	Cardiovascular devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F11	Non-Serious Injury/ Illness/ Impairment	A070504	Premature Discharge of Battery		
					CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F11	Non-Serious Injury/ Illness/ Impairment	A070504	Premature Discharge of Battery		
					CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F11	Non-Serious Injury/ Illness/ Impairment	A070504	Premature Discharge of Battery		
					CT233	Surgical	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F11	Non-Serious Injury/ Illness/ Impairment	A070504	Premature Discharge of Battery		
					CT2406	Implantable	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F11	Non-Serious Injury/ Illness/ Impairment	A070504	Premature Discharge of Battery		
					CT2407	Active implantable	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F11	Non-Serious Injury/ Illness/ Impairment	A070504	Premature Discharge of Battery		
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F11	Non-Serious Injury/ Illness/ Impairment	A070504	Premature Discharge of Battery		
					CT2727	Active implant devices/ancillaries	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F11	Non-Serious Injury/ Illness/ Impairment	A070504	Premature Discharge of Battery		
					CT2750	Implantable cardiac electrical stimulators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F11	Non-Serious Injury/ Illness/ Impairment	A070504	Premature Discharge of Battery		
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F11	Non-Serious Injury/ Illness/ Impairment	A070504	Premature Discharge of Battery		
					CT2985	Active	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F11	Non-Serious Injury/ Illness/ Impairment	A070504	Premature Discharge of Battery		
					CT3001	Implantable metal	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F11	Non-Serious Injury/ Illness/ Impairment	A070504	Premature Discharge of Battery		
					CT3002	Implantable synthetic polymer	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F11	Non-Serious Injury/ Illness/ Impairment	A070504	Premature Discharge of Battery		
					CT3006	Implantable inorganic material	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F11	Non-Serious Injury/ Illness/ Impairment	A070504	Premature Discharge of Battery		
					CT314	Electrically powered	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F11	Non-Serious Injury/ Illness/ Impairment	A070504	Premature Discharge of Battery		
					CT321	Long-term surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F11	Non-Serious Injury/ Illness/ Impairment	A070504	Premature Discharge of Battery		

					CT449	Generators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F11	Non-Serious Injury/ Illness/ Impairment	A070504	Premature Discharge of Battery		
					CT981	Single use	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F11	Non-Serious Injury/ Illness/ Impairment	A070504	Premature Discharge of Battery		
					CT983	Surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F11	Non-Serious Injury/ Illness/ Impairment	A070504	Premature Discharge of Battery		
11110	2021	All other reportable incidents	47265	Dual-chamber implantable pacemaker, rate-responsive	CT141	Electrodes/Leads and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	19/05/2021	19/05/2021
							F1901	Additional Surgery	A070504	Premature Discharge of Battery	19/05/2021	19/05/2021
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	19/05/2021	19/05/2021
					CT145	Cardiovascular devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	19/05/2021	19/05/2021
							F1901	Additional Surgery	A070504	Premature Discharge of Battery	19/05/2021	19/05/2021
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	19/05/2021	19/05/2021
					CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	19/05/2021	19/05/2021
							F1901	Additional Surgery	A070504	Premature Discharge of Battery	19/05/2021	19/05/2021
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	19/05/2021	19/05/2021
					CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	19/05/2021	19/05/2021
							F1901	Additional Surgery	A070504	Premature Discharge of Battery	19/05/2021	19/05/2021
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	19/05/2021	19/05/2021
					CT233	Surgical	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	19/05/2021	19/05/2021
							F1901	Additional Surgery	A070504	Premature Discharge of Battery	19/05/2021	19/05/2021
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	19/05/2021	19/05/2021
					CT2406	Implantable	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	19/05/2021	19/05/2021
							F1901	Additional Surgery	A070504	Premature Discharge of Battery	19/05/2021	19/05/2021
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	19/05/2021	19/05/2021
					CT2407	Active implantable	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	19/05/2021	19/05/2021
							F1901	Additional Surgery	A070504	Premature Discharge of Battery	19/05/2021	19/05/2021
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	19/05/2021	19/05/2021
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	19/05/2021	19/05/2021
							F1901	Additional Surgery	A070504	Premature Discharge of Battery	19/05/2021	19/05/2021
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	19/05/2021	19/05/2021
					CT2727	Active implant devices/ancillaries	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	19/05/2021	19/05/2021
							F1901	Additional Surgery	A070504	Premature Discharge of Battery	19/05/2021	19/05/2021
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	19/05/2021	19/05/2021
					CT2750	Implantable cardiac electrical stimulators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	19/05/2021	19/05/2021
							F1901	Additional Surgery	A070504	Premature Discharge of Battery	19/05/2021	19/05/2021
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	19/05/2021	19/05/2021
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	19/05/2021	19/05/2021
							F1901	Additional Surgery	A070504	Premature Discharge of Battery	19/05/2021	19/05/2021
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	19/05/2021	19/05/2021
					CT2985	Active	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	19/05/2021	19/05/2021
							F1901	Additional Surgery	A070504	Premature Discharge of Battery	19/05/2021	19/05/2021
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	19/05/2021	19/05/2021
					CT3001	Implantable metal	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	19/05/2021	19/05/2021
							F1901	Additional Surgery	A070504	Premature Discharge of Battery	19/05/2021	19/05/2021
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	19/05/2021	19/05/2021
					CT3002	Implantable synthetic polymer	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	19/05/2021	19/05/2021
							F1901	Additional Surgery	A070504	Premature Discharge of Battery	19/05/2021	19/05/2021
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	19/05/2021	19/05/2021
					CT3006	Implantable inorganic material	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	19/05/2021	19/05/2021
							F1901	Additional Surgery	A070504	Premature Discharge of Battery	19/05/2021	19/05/2021
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	19/05/2021	19/05/2021
					CT314	Electrically powered	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	19/05/2021	19/05/2021
							F1901	Additional Surgery	A070504	Premature Discharge of Battery	19/05/2021	19/05/2021
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	19/05/2021	19/05/2021
					CT321	Long-term surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	19/05/2021	19/05/2021
							F1901	Additional Surgery	A070504	Premature Discharge of Battery	19/05/2021	19/05/2021
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	19/05/2021	19/05/2021
					CT449	Generators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	19/05/2021	19/05/2021
							F1901	Additional Surgery	A070504	Premature Discharge of Battery	19/05/2021	19/05/2021
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	19/05/2021	19/05/2021
					CT981	Single use	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	19/05/2021	19/05/2021
							F1901	Additional Surgery	A070504	Premature Discharge of Battery	19/05/2021	19/05/2021
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	19/05/2021	19/05/2021
					CT983	Surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	19/05/2021	19/05/2021
							F1901	Additional Surgery	A070504	Premature Discharge of Battery	19/05/2021	19/05/2021
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	19/05/2021	19/05/2021
11111	2021	All other reportable incidents	47263	Cardiac resynchronization therapy implantable pacemaker	CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A0908	Incorrect, Inadequate or Imprecise Result or Readings	11/05/2021	11/05/2021
							F08	Hospitalization or Prolonged Hospitalization	A0908	Incorrect, Inadequate or Imprecise Result or Readings	11/05/2021	11/05/2021
							F1905	Device Revision or Replacement	A0908	Incorrect, Inadequate or Imprecise Result or Readings	11/05/2021	11/05/2021
					CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A0908	Incorrect, Inadequate or Imprecise Result or Readings	11/05/2021	11/05/2021
							F08	Hospitalization or Prolonged Hospitalization	A0908	Incorrect, Inadequate or Imprecise Result or Readings	11/05/2021	11/05/2021
							F1905	Device Revision or Replacement	A0908	Incorrect, Inadequate or Imprecise Result or Readings	11/05/2021	11/05/2021
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A0908	Incorrect, Inadequate or Imprecise Result or Readings	11/05/2021	11/05/2021
							F08	Hospitalization or Prolonged Hospitalization	A0908	Incorrect, Inadequate or Imprecise Result or Readings	11/05/2021	11/05/2021
							F1905	Device Revision or Replacement	A0908	Incorrect, Inadequate or Imprecise Result or Readings	11/05/2021	11/05/2021
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A0908	Incorrect, Inadequate or Imprecise Result or Readings	11/05/2021	11/05/2021
							F08	Hospitalization or Prolonged Hospitalization	A0908	Incorrect, Inadequate or Imprecise Result or Readings	11/05/2021	11/05/2021
							F1905	Device Revision or Replacement	A0908	Incorrect, Inadequate or Imprecise Result or Readings	11/05/2021	11/05/2021
11112	2021	All other reportable incidents	17882	Physiologic-monitoring defibrillation system	CT1684	Cardiac pacemakers and associated devices	E2401	Insufficient Information	A071301	Failure to Deliver Shock/Stimulation		

							E2403	No Clinical Signs, Symptoms or Conditions	A071301	Failure to Deliver Shock/Stimulation		
							F24	Insufficient Health Impact Information	A071301	Failure to Deliver Shock/Stimulation		
11113	2021	All other reportable incidents	17882	Physiologic-monitoring defibrillation system	CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A07	Electrical /Electronic Property Problem		
							F27	Problem Identified During Non-Clinical Procedure	A07	Electrical /Electronic Property Problem		
11114	2021	All other reportable incidents	35223	Endocardial/interventricular septal pacing lead	CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A051201	Device Dislodged or Dislocated	14/05/2021	14/05/2021
									A070102	High Capture Threshold	14/05/2021	14/05/2021
									A070901	Decreased Sensitivity	14/05/2021	14/05/2021
									A070909	Oversensing	14/05/2021	14/05/2021
									A0712	Pacing Problem	14/05/2021	14/05/2021
									A0722	Impedance Problem	14/05/2021	14/05/2021
							F26	No Health Consequences or Impact	A051201	Device Dislodged or Dislocated	14/05/2021	14/05/2021
									A070102	High Capture Threshold	14/05/2021	14/05/2021
									A070901	Decreased Sensitivity	14/05/2021	14/05/2021
									A070909	Oversensing	14/05/2021	14/05/2021
									A0712	Pacing Problem	14/05/2021	14/05/2021
									A0722	Impedance Problem	14/05/2021	14/05/2021
11115	2021	All other reportable incidents	47265	Dual-chamber implantable pacemaker, rate-responsive	CT141	Electrodes/Leads and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	19/05/2021	19/05/2021
							F1901	Additional Surgery	A070504	Premature Discharge of Battery	19/05/2021	19/05/2021
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	19/05/2021	19/05/2021
					CT145	Cardiovascular devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	19/05/2021	19/05/2021
							F1901	Additional Surgery	A070504	Premature Discharge of Battery	19/05/2021	19/05/2021
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	19/05/2021	19/05/2021
					CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	19/05/2021	19/05/2021
							F1901	Additional Surgery	A070504	Premature Discharge of Battery	19/05/2021	19/05/2021
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	19/05/2021	19/05/2021
					CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	19/05/2021	19/05/2021
							F1901	Additional Surgery	A070504	Premature Discharge of Battery	19/05/2021	19/05/2021
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	19/05/2021	19/05/2021
					CT233	Surgical	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	19/05/2021	19/05/2021
							F1901	Additional Surgery	A070504	Premature Discharge of Battery	19/05/2021	19/05/2021
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	19/05/2021	19/05/2021
					CT2406	Implantable	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	19/05/2021	19/05/2021
							F1901	Additional Surgery	A070504	Premature Discharge of Battery	19/05/2021	19/05/2021
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	19/05/2021	19/05/2021
					CT2407	Active implantable	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	19/05/2021	19/05/2021
							F1901	Additional Surgery	A070504	Premature Discharge of Battery	19/05/2021	19/05/2021
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	19/05/2021	19/05/2021
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	19/05/2021	19/05/2021
							F1901	Additional Surgery	A070504	Premature Discharge of Battery	19/05/2021	19/05/2021
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	19/05/2021	19/05/2021
					CT2727	Active implant devices/ancillaries	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	19/05/2021	19/05/2021
							F1901	Additional Surgery	A070504	Premature Discharge of Battery	19/05/2021	19/05/2021
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	19/05/2021	19/05/2021
					CT2750	Implantable cardiac electrical stimulators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	19/05/2021	19/05/2021
							F1901	Additional Surgery	A070504	Premature Discharge of Battery	19/05/2021	19/05/2021
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	19/05/2021	19/05/2021
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	19/05/2021	19/05/2021
							F1901	Additional Surgery	A070504	Premature Discharge of Battery	19/05/2021	19/05/2021
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	19/05/2021	19/05/2021
					CT2985	Active	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	19/05/2021	19/05/2021
							F1901	Additional Surgery	A070504	Premature Discharge of Battery	19/05/2021	19/05/2021
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	19/05/2021	19/05/2021
					CT3001	Implantable metal	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	19/05/2021	19/05/2021
							F1901	Additional Surgery	A070504	Premature Discharge of Battery	19/05/2021	19/05/2021
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	19/05/2021	19/05/2021
					CT3002	Implantable synthetic polymer	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	19/05/2021	19/05/2021
							F1901	Additional Surgery	A070504	Premature Discharge of Battery	19/05/2021	19/05/2021
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	19/05/2021	19/05/2021
					CT3006	Implantable inorganic material	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	19/05/2021	19/05/2021
							F1901	Additional Surgery	A070504	Premature Discharge of Battery	19/05/2021	19/05/2021
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	19/05/2021	19/05/2021
					CT314	Electrically powered	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	19/05/2021	19/05/2021
							F1901	Additional Surgery	A070504	Premature Discharge of Battery	19/05/2021	19/05/2021
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	19/05/2021	19/05/2021
					CT321	Long-term surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	19/05/2021	19/05/2021
							F1901	Additional Surgery	A070504	Premature Discharge of Battery	19/05/2021	19/05/2021
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	19/05/2021	19/05/2021
					CT449	Generators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	19/05/2021	19/05/2021
							F1901	Additional Surgery	A070504	Premature Discharge of Battery	19/05/2021	19/05/2021
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	19/05/2021	19/05/2021
					CT981	Single use	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	19/05/2021	19/05/2021
							F1901	Additional Surgery	A070504	Premature Discharge of Battery	19/05/2021	19/05/2021
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	19/05/2021	19/05/2021
					CT983	Surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	19/05/2021	19/05/2021
							F1901	Additional Surgery	A070504	Premature Discharge of Battery	19/05/2021	19/05/2021
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	19/05/2021	19/05/2021
11116	2021	Death	47270	Cardiac resynchronization therapy implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E0602	Cardiac Arrest	A070101	Failure to Capture		
									A070909	Oversensing		
									A070910	Undersensing		
									A071206	Pacing Asynchronously		
							E0611	Heart Failure/Congestive Heart Failure	A070101	Failure to Capture		
									A070909	Oversensing		
									A070910	Undersensing		
									A071206	Pacing Asynchronously		
							F02	Death	A070101	Failure to Capture		
									A070909	Oversensing		
									A070910	Undersensing		
									A071206	Pacing Asynchronously		

11117	2021	All other reportable incidents	35223	Endocardial/interventricular septal pacing lead	CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070909	Oversensing		
							F26	No Health Consequences or Impact	A070909	Oversensing		
11118	2021	All other reportable incidents	47270	Cardiac resynchronization therapy implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	24/05/2021	24/05/2021
							F1901	Additional Surgery	A070504	Premature Discharge of Battery	24/05/2021	24/05/2021
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	24/05/2021	24/05/2021
11119	2021	Death	35223	Endocardial/interventricular septal pacing lead	CT1684	Cardiac pacemakers and associated devices	E0515	Vascular Dissection	A070101	Failure to Capture	20/05/2021	20/05/2021
									A070102	High Capture Threshold	20/05/2021	20/05/2021
									A0722	Impedance Problem	20/05/2021	20/05/2021
									A072201	High Impedance	20/05/2021	20/05/2021
							E060106	Heart Block	A070101	Failure to Capture	20/05/2021	20/05/2021
									A070102	High Capture Threshold	20/05/2021	20/05/2021
									A0722	Impedance Problem	20/05/2021	20/05/2021
									A072201	High Impedance	20/05/2021	20/05/2021
							E2114	Perforation	A070101	Failure to Capture	20/05/2021	20/05/2021
									A070102	High Capture Threshold	20/05/2021	20/05/2021
									A0722	Impedance Problem	20/05/2021	20/05/2021
									A072201	High Impedance	20/05/2021	20/05/2021
							E2401	Insufficient Information	A070101	Failure to Capture	20/05/2021	20/05/2021
									A070102	High Capture Threshold	20/05/2021	20/05/2021
									A0722	Impedance Problem	20/05/2021	20/05/2021
									A072201	High Impedance	20/05/2021	20/05/2021
							F02	Death	A070101	Failure to Capture	20/05/2021	20/05/2021
									A070102	High Capture Threshold	20/05/2021	20/05/2021
									A0722	Impedance Problem	20/05/2021	20/05/2021
									A072201	High Impedance	20/05/2021	20/05/2021
							F1203	Life Threatening Illness or Injury	A070101	Failure to Capture	20/05/2021	20/05/2021
									A070102	High Capture Threshold	20/05/2021	20/05/2021
									A0722	Impedance Problem	20/05/2021	20/05/2021
									A072201	High Impedance	20/05/2021	20/05/2021
							F14	Prolonged Episode of Care	A070101	Failure to Capture	20/05/2021	20/05/2021
									A070102	High Capture Threshold	20/05/2021	20/05/2021
									A0722	Impedance Problem	20/05/2021	20/05/2021
									A072201	High Impedance	20/05/2021	20/05/2021
							F1901	Additional Surgery	A070101	Failure to Capture	20/05/2021	20/05/2021
									A070102	High Capture Threshold	20/05/2021	20/05/2021
									A0722	Impedance Problem	20/05/2021	20/05/2021
									A072201	High Impedance	20/05/2021	20/05/2021
							F1905	Device Revision or Replacement	A070101	Failure to Capture	20/05/2021	20/05/2021
									A070102	High Capture Threshold	20/05/2021	20/05/2021
									A0722	Impedance Problem	20/05/2021	20/05/2021
									A072201	High Impedance	20/05/2021	20/05/2021
11120	2021	All other reportable incidents	47205	Cardiac pulse generator programmer	CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A0701	Capturing Problem		
									A070101	Failure to Capture		
							F26	No Health Consequences or Impact	A0701	Capturing Problem		
									A070101	Failure to Capture		
11121	2021	All other reportable incidents	47265	Dual-chamber implantable pacemaker, rate-responsive	CT141	Electrodes/Leads and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	28/06/2021	28/06/2021
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	28/06/2021	28/06/2021
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	28/06/2021	28/06/2021
					CT145	Cardiovascular devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	28/06/2021	28/06/2021
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	28/06/2021	28/06/2021
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	28/06/2021	28/06/2021
					CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	28/06/2021	28/06/2021
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	28/06/2021	28/06/2021
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	28/06/2021	28/06/2021
					CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	28/06/2021	28/06/2021
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	28/06/2021	28/06/2021
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	28/06/2021	28/06/2021
					CT233	Surgical	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	28/06/2021	28/06/2021
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	28/06/2021	28/06/2021
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	28/06/2021	28/06/2021
					CT2406	Implantable	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	28/06/2021	28/06/2021
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	28/06/2021	28/06/2021
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	28/06/2021	28/06/2021
					CT2407	Active implantable	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	28/06/2021	28/06/2021
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	28/06/2021	28/06/2021
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	28/06/2021	28/06/2021
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	28/06/2021	28/06/2021
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	28/06/2021	28/06/2021
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	28/06/2021	28/06/2021
					CT2727	Active implant devices/ancillaries	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	28/06/2021	28/06/2021
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	28/06/2021	28/06/2021
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	28/06/2021	28/06/2021
					CT2750	Implantable cardiac electrical stimulators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	28/06/2021	28/06/2021
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	28/06/2021	28/06/2021
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	28/06/2021	28/06/2021
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	28/06/2021	28/06/2021
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	28/06/2021	28/06/2021
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	28/06/2021	28/06/2021
					CT2985	Active	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	28/06/2021	28/06/2021
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	28/06/2021	28/06/2021
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	28/06/2021	28/06/2021
					CT3001	Implantable metal	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	28/06/2021	28/06/2021
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	28/06/2021	28/06/2021
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	28/06/2021	28/06/2021
					CT3002	Implantable synthetic polymer	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	28/06/2021	28/06/2021
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	28/06/2021	28/06/2021
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	28/06/2021	28/06/2021
					CT3006	Implantable inorganic material	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	28/06/2021	28/06/2021

							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	28/06/2021	28/06/2021
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	28/06/2021	28/06/2021
					CT314	Electrically powered	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	28/06/2021	28/06/2021
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	28/06/2021	28/06/2021
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	28/06/2021	28/06/2021
					CT321	Long-term surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	28/06/2021	28/06/2021
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	28/06/2021	28/06/2021
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	28/06/2021	28/06/2021
					CT449	Generators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	28/06/2021	28/06/2021
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	28/06/2021	28/06/2021
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	28/06/2021	28/06/2021
					CT981	Single use	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	28/06/2021	28/06/2021
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	28/06/2021	28/06/2021
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	28/06/2021	28/06/2021
					CT983	Surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	28/06/2021	28/06/2021
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	28/06/2021	28/06/2021
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	28/06/2021	28/06/2021
11122	2021	All other reportable incidents	47265	Dual-chamber implantable pacemaker, rate-responsive	CT141	Electrodes/Leads and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	22/06/2021	22/06/2021
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	22/06/2021	22/06/2021
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	22/06/2021	22/06/2021
					CT145	Cardiovascular devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	22/06/2021	22/06/2021
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	22/06/2021	22/06/2021
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	22/06/2021	22/06/2021
					CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	22/06/2021	22/06/2021
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	22/06/2021	22/06/2021
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	22/06/2021	22/06/2021
					CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	22/06/2021	22/06/2021
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	22/06/2021	22/06/2021
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	22/06/2021	22/06/2021
					CT233	Surgical	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	22/06/2021	22/06/2021
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	22/06/2021	22/06/2021
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	22/06/2021	22/06/2021
					CT2406	Implantable	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	22/06/2021	22/06/2021
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	22/06/2021	22/06/2021
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	22/06/2021	22/06/2021
					CT2407	Active implantable	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	22/06/2021	22/06/2021
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	22/06/2021	22/06/2021
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	22/06/2021	22/06/2021
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	22/06/2021	22/06/2021
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	22/06/2021	22/06/2021
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	22/06/2021	22/06/2021
					CT2727	Active implant devices/ancillaries	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	22/06/2021	22/06/2021
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	22/06/2021	22/06/2021
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	22/06/2021	22/06/2021
					CT2750	Implantable cardiac electrical stimulators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	22/06/2021	22/06/2021
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	22/06/2021	22/06/2021
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	22/06/2021	22/06/2021
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	22/06/2021	22/06/2021
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	22/06/2021	22/06/2021
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	22/06/2021	22/06/2021
					CT2985	Active	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	22/06/2021	22/06/2021
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	22/06/2021	22/06/2021
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	22/06/2021	22/06/2021
					CT3001	Implantable metal	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	22/06/2021	22/06/2021
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	22/06/2021	22/06/2021
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	22/06/2021	22/06/2021
					CT3002	Implantable synthetic polymer	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	22/06/2021	22/06/2021
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	22/06/2021	22/06/2021
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	22/06/2021	22/06/2021
					CT3006	Implantable inorganic material	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	22/06/2021	22/06/2021
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	22/06/2021	22/06/2021
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	22/06/2021	22/06/2021
					CT314	Electrically powered	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	22/06/2021	22/06/2021
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	22/06/2021	22/06/2021
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	22/06/2021	22/06/2021
					CT321	Long-term surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	22/06/2021	22/06/2021
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	22/06/2021	22/06/2021
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	22/06/2021	22/06/2021
					CT449	Generators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	22/06/2021	22/06/2021
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	22/06/2021	22/06/2021
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	22/06/2021	22/06/2021
					CT981	Single use	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	22/06/2021	22/06/2021
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	22/06/2021	22/06/2021
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	22/06/2021	22/06/2021
					CT983	Surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	22/06/2021	22/06/2021
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	22/06/2021	22/06/2021
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	22/06/2021	22/06/2021
11123	2021	All other reportable incidents	47265	Dual-chamber implantable pacemaker, rate-responsive	CT141	Electrodes/Leads and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	24/05/2021	24/05/2021
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	24/05/2021	24/05/2021
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	24/05/2021	24/05/2021
					CT145	Cardiovascular devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	24/05/2021	24/05/2021
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	24/05/2021	24/05/2021
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	24/05/2021	24/05/2021
					CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	24/05/2021	24/05/2021
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	24/05/2021	24/05/2021
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	24/05/2021	24/05/2021
					CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	24/05/2021	24/05/2021
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	24/05/2021	24/05/2021

							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	24/05/2021	24/05/2021
					CT233	Surgical	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	24/05/2021	24/05/2021
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	24/05/2021	24/05/2021
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	24/05/2021	24/05/2021
					CT2406	Implantable	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	24/05/2021	24/05/2021
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	24/05/2021	24/05/2021
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	24/05/2021	24/05/2021
					CT2407	Active implantable	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	24/05/2021	24/05/2021
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	24/05/2021	24/05/2021
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	24/05/2021	24/05/2021
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	24/05/2021	24/05/2021
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	24/05/2021	24/05/2021
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	24/05/2021	24/05/2021
					CT2727	Active implant devices/ancillaries	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	24/05/2021	24/05/2021
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	24/05/2021	24/05/2021
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	24/05/2021	24/05/2021
					CT2750	Implantable cardiac electrical stimulators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	24/05/2021	24/05/2021
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	24/05/2021	24/05/2021
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	24/05/2021	24/05/2021
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	24/05/2021	24/05/2021
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	24/05/2021	24/05/2021
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	24/05/2021	24/05/2021
					CT2985	Active	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	24/05/2021	24/05/2021
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	24/05/2021	24/05/2021
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	24/05/2021	24/05/2021
					CT3001	Implantable metal	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	24/05/2021	24/05/2021
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	24/05/2021	24/05/2021
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	24/05/2021	24/05/2021
					CT3002	Implantable synthetic polymer	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	24/05/2021	24/05/2021
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	24/05/2021	24/05/2021
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	24/05/2021	24/05/2021
					CT3006	Implantable inorganic material	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	24/05/2021	24/05/2021
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	24/05/2021	24/05/2021
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	24/05/2021	24/05/2021
					CT314	Electrically powered	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	24/05/2021	24/05/2021
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	24/05/2021	24/05/2021
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	24/05/2021	24/05/2021
					CT321	Long-term surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	24/05/2021	24/05/2021
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	24/05/2021	24/05/2021
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	24/05/2021	24/05/2021
					CT449	Generators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	24/05/2021	24/05/2021
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	24/05/2021	24/05/2021
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	24/05/2021	24/05/2021
					CT981	Single use	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	24/05/2021	24/05/2021
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	24/05/2021	24/05/2021
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	24/05/2021	24/05/2021
					CT983	Surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	24/05/2021	24/05/2021
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	24/05/2021	24/05/2021
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	24/05/2021	24/05/2021
11124	2021	All other reportable incidents	47265	Dual-chamber implantable pacemaker, rate-responsive	CT141	Electrodes/Leads and associated devices	E060104	Bradycardia	A070909	Oversensing		
									A0712	Pacing Problem		
									A072201	High Impedance		
									A090801	Signal Artifact/Noise		
							E2403	No Clinical Signs, Symptoms or Conditions	A070909	Oversensing		
									A0712	Pacing Problem		
									A072201	High Impedance		
									A090801	Signal Artifact/Noise		
							F15	Recognized Device or Procedural Complication	A070909	Oversensing		
									A0712	Pacing Problem		
									A072201	High Impedance		
									A090801	Signal Artifact/Noise		
					CT145	Cardiovascular devices	E060104	Bradycardia	A070909	Oversensing		
									A0712	Pacing Problem		
									A072201	High Impedance		
									A090801	Signal Artifact/Noise		
							E2403	No Clinical Signs, Symptoms or Conditions	A070909	Oversensing		
									A0712	Pacing Problem		
									A072201	High Impedance		
									A090801	Signal Artifact/Noise		
							F15	Recognized Device or Procedural Complication	A070909	Oversensing		
									A0712	Pacing Problem		
									A072201	High Impedance		
									A090801	Signal Artifact/Noise		
					CT1684	Cardiac pacemakers and associated devices	E060104	Bradycardia	A070909	Oversensing		
									A0712	Pacing Problem		
									A072201	High Impedance		
									A090801	Signal Artifact/Noise		
							E2403	No Clinical Signs, Symptoms or Conditions	A070909	Oversensing		
									A0712	Pacing Problem		
									A072201	High Impedance		
									A090801	Signal Artifact/Noise		
							F15	Recognized Device or Procedural Complication	A070909	Oversensing		
									A0712	Pacing Problem		
									A072201	High Impedance		
									A090801	Signal Artifact/Noise		
					CT1685	Cardiac pacemakers	E060104	Bradycardia	A070909	Oversensing		
									A0712	Pacing Problem		
									A072201	High Impedance		

									A090801	Signal Artifact/Noise		
							E2403	No Clinical Signs, Symptoms or Conditions	A070909	Oversensing		
									A0712	Pacing Problem		
									A072201	High Impedance		
									A090801	Signal Artifact/Noise		
							F15	Recognized Device or Procedural Complication	A070909	Oversensing		
									A0712	Pacing Problem		
									A072201	High Impedance		
									A090801	Signal Artifact/Noise		
					CT233	Surgical	E060104	Bradycardia	A070909	Oversensing		
									A0712	Pacing Problem		
									A072201	High Impedance		
									A090801	Signal Artifact/Noise		
							E2403	No Clinical Signs, Symptoms or Conditions	A070909	Oversensing		
									A0712	Pacing Problem		
									A072201	High Impedance		
									A090801	Signal Artifact/Noise		
							F15	Recognized Device or Procedural Complication	A070909	Oversensing		
									A0712	Pacing Problem		
									A072201	High Impedance		
									A090801	Signal Artifact/Noise		
					CT2406	Implantable	E060104	Bradycardia	A070909	Oversensing		
									A0712	Pacing Problem		
									A072201	High Impedance		
									A090801	Signal Artifact/Noise		
							E2403	No Clinical Signs, Symptoms or Conditions	A070909	Oversensing		
									A0712	Pacing Problem		
									A072201	High Impedance		
									A090801	Signal Artifact/Noise		
							F15	Recognized Device or Procedural Complication	A070909	Oversensing		
									A0712	Pacing Problem		
									A072201	High Impedance		
									A090801	Signal Artifact/Noise		
					CT2407	Active implantable	E060104	Bradycardia	A070909	Oversensing		
									A0712	Pacing Problem		
									A072201	High Impedance		
									A090801	Signal Artifact/Noise		
							E2403	No Clinical Signs, Symptoms or Conditions	A070909	Oversensing		
									A0712	Pacing Problem		
									A072201	High Impedance		
									A090801	Signal Artifact/Noise		
							F15	Recognized Device or Procedural Complication	A070909	Oversensing		
									A0712	Pacing Problem		
									A072201	High Impedance		
									A090801	Signal Artifact/Noise		
					CT257	Implantable cardiac pacemakers	E060104	Bradycardia	A070909	Oversensing		
									A0712	Pacing Problem		
									A072201	High Impedance		
									A090801	Signal Artifact/Noise		
							E2403	No Clinical Signs, Symptoms or Conditions	A070909	Oversensing		
									A0712	Pacing Problem		
									A072201	High Impedance		
									A090801	Signal Artifact/Noise		
							F15	Recognized Device or Procedural Complication	A070909	Oversensing		
									A0712	Pacing Problem		
									A072201	High Impedance		
									A090801	Signal Artifact/Noise		
					CT2727	Active implant devices/ancillaries	E060104	Bradycardia	A070909	Oversensing		
									A0712	Pacing Problem		
									A072201	High Impedance		
									A090801	Signal Artifact/Noise		
							E2403	No Clinical Signs, Symptoms or Conditions	A070909	Oversensing		
									A0712	Pacing Problem		
									A072201	High Impedance		
									A090801	Signal Artifact/Noise		
							F15	Recognized Device or Procedural Complication	A070909	Oversensing		
									A0712	Pacing Problem		
									A072201	High Impedance		
									A090801	Signal Artifact/Noise		
					CT2750	Implantable cardiac electrical stimulators and associated devices	E060104	Bradycardia	A070909	Oversensing		
									A0712	Pacing Problem		
									A072201	High Impedance		
									A090801	Signal Artifact/Noise		
							E2403	No Clinical Signs, Symptoms or Conditions	A070909	Oversensing		
									A0712	Pacing Problem		
									A072201	High Impedance		
									A090801	Signal Artifact/Noise		
							F15	Recognized Device or Procedural Complication	A070909	Oversensing		
									A0712	Pacing Problem		
									A072201	High Impedance		
									A090801	Signal Artifact/Noise		
					CT2866	Implantable cardiac electrical stimulators	E060104	Bradycardia	A070909	Oversensing		
									A0712	Pacing Problem		
									A072201	High Impedance		
									A090801	Signal Artifact/Noise		
							E2403	No Clinical Signs, Symptoms or Conditions	A070909	Oversensing		
									A0712	Pacing Problem		
									A072201	High Impedance		

									A090801	Signal Artifact/Noise		
							F15	Recognized Device or Procedural Complication	A070909	Oversensing		
									A0712	Pacing Problem		
									A072201	High Impedance		
									A090801	Signal Artifact/Noise		
					CT2985	Active	E060104	Bradycardia	A070909	Oversensing		
									A0712	Pacing Problem		
									A072201	High Impedance		
									A090801	Signal Artifact/Noise		
							E2403	No Clinical Signs, Symptoms or Conditions	A070909	Oversensing		
									A0712	Pacing Problem		
									A072201	High Impedance		
									A090801	Signal Artifact/Noise		
							F15	Recognized Device or Procedural Complication	A070909	Oversensing		
									A0712	Pacing Problem		
									A072201	High Impedance		
									A090801	Signal Artifact/Noise		
					CT3001	Implantable metal	E060104	Bradycardia	A070909	Oversensing		
									A0712	Pacing Problem		
									A072201	High Impedance		
									A090801	Signal Artifact/Noise		
							E2403	No Clinical Signs, Symptoms or Conditions	A070909	Oversensing		
									A0712	Pacing Problem		
									A072201	High Impedance		
									A090801	Signal Artifact/Noise		
							F15	Recognized Device or Procedural Complication	A070909	Oversensing		
									A0712	Pacing Problem		
									A072201	High Impedance		
									A090801	Signal Artifact/Noise		
					CT3002	Implantable synthetic polymer	E060104	Bradycardia	A070909	Oversensing		
									A0712	Pacing Problem		
									A072201	High Impedance		
									A090801	Signal Artifact/Noise		
							E2403	No Clinical Signs, Symptoms or Conditions	A070909	Oversensing		
									A0712	Pacing Problem		
									A072201	High Impedance		
									A090801	Signal Artifact/Noise		
							F15	Recognized Device or Procedural Complication	A070909	Oversensing		
									A0712	Pacing Problem		
									A072201	High Impedance		
									A090801	Signal Artifact/Noise		
					CT3006	Implantable inorganic material	E060104	Bradycardia	A070909	Oversensing		
									A0712	Pacing Problem		
									A072201	High Impedance		
									A090801	Signal Artifact/Noise		
							E2403	No Clinical Signs, Symptoms or Conditions	A070909	Oversensing		
									A0712	Pacing Problem		
									A072201	High Impedance		
									A090801	Signal Artifact/Noise		
							F15	Recognized Device or Procedural Complication	A070909	Oversensing		
									A0712	Pacing Problem		
									A072201	High Impedance		
									A090801	Signal Artifact/Noise		
					CT314	Electrically powered	E060104	Bradycardia	A070909	Oversensing		
									A0712	Pacing Problem		
									A072201	High Impedance		
									A090801	Signal Artifact/Noise		
							E2403	No Clinical Signs, Symptoms or Conditions	A070909	Oversensing		
									A0712	Pacing Problem		
									A072201	High Impedance		
									A090801	Signal Artifact/Noise		
							F15	Recognized Device or Procedural Complication	A070909	Oversensing		
									A0712	Pacing Problem		
									A072201	High Impedance		
									A090801	Signal Artifact/Noise		
					CT321	Long-term surgical invasive	E060104	Bradycardia	A070909	Oversensing		
									A0712	Pacing Problem		
									A072201	High Impedance		
									A090801	Signal Artifact/Noise		
							E2403	No Clinical Signs, Symptoms or Conditions	A070909	Oversensing		
									A0712	Pacing Problem		
									A072201	High Impedance		
									A090801	Signal Artifact/Noise		
							F15	Recognized Device or Procedural Complication	A070909	Oversensing		
									A0712	Pacing Problem		
									A072201	High Impedance		
									A090801	Signal Artifact/Noise		
					CT449	Generators and associated devices	E060104	Bradycardia	A070909	Oversensing		
									A0712	Pacing Problem		
									A072201	High Impedance		
									A090801	Signal Artifact/Noise		
							E2403	No Clinical Signs, Symptoms or Conditions	A070909	Oversensing		
									A0712	Pacing Problem		
									A072201	High Impedance		
									A090801	Signal Artifact/Noise		
							F15	Recognized Device or Procedural Complication	A070909	Oversensing		
									A0712	Pacing Problem		
									A072201	High Impedance		

									A090801	Signal Artifact/Noise		
					CT981	Single use	E060104	Bradycardia	A070909	Oversensing		
									A0712	Pacing Problem		
									A072201	High Impedance		
									A090801	Signal Artifact/Noise		
							E2403	No Clinical Signs, Symptoms or Conditions	A070909	Oversensing		
									A0712	Pacing Problem		
									A072201	High Impedance		
									A090801	Signal Artifact/Noise		
							F15	Recognized Device or Procedural Complication	A070909	Oversensing		
									A0712	Pacing Problem		
									A072201	High Impedance		
									A090801	Signal Artifact/Noise		
					CT983	Surgical invasive	E060104	Bradycardia	A070909	Oversensing		
									A0712	Pacing Problem		
									A072201	High Impedance		
									A090801	Signal Artifact/Noise		
							E2403	No Clinical Signs, Symptoms or Conditions	A070909	Oversensing		
									A0712	Pacing Problem		
									A072201	High Impedance		
									A090801	Signal Artifact/Noise		
							F15	Recognized Device or Procedural Complication	A070909	Oversensing		
									A0712	Pacing Problem		
									A072201	High Impedance		
									A090801	Signal Artifact/Noise		
11125	2021	All other reportable incidents	36241	Pacemaker/defibrillator lead	CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A05	Mechanical Problem	20/05/2021	20/05/2021
									A072201	High Impedance	20/05/2021	20/05/2021
							F1908	Prolonged Surgery	A05	Mechanical Problem	20/05/2021	20/05/2021
									A072201	High Impedance	20/05/2021	20/05/2021
11126	2021	All other reportable incidents	35852	Single-chamber implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem	29/06/2021	29/06/2021
									A070504	Premature Discharge of Battery	29/06/2021	29/06/2021
							F08	Hospitalization or Prolonged Hospitalization	A0705	Battery Problem	29/06/2021	29/06/2021
									A070504	Premature Discharge of Battery	29/06/2021	29/06/2021
							F1903	Device Explantation	A0705	Battery Problem	29/06/2021	29/06/2021
									A070504	Premature Discharge of Battery	29/06/2021	29/06/2021
11127	2021	All other reportable incidents	36241	Pacemaker/defibrillator lead	CT1684	Cardiac pacemakers and associated devices	E020201	Anxiety	A0203	Defective Device	26/05/2021	26/05/2021
									A070909	Oversensing	26/05/2021	26/05/2021
									A0712	Pacing Problem	26/05/2021	26/05/2021
									A072201	High Impedance	26/05/2021	26/05/2021
							F08	Hospitalization or Prolonged Hospitalization	A0203	Defective Device	26/05/2021	26/05/2021
									A070909	Oversensing	26/05/2021	26/05/2021
									A0712	Pacing Problem	26/05/2021	26/05/2021
									A072201	High Impedance	26/05/2021	26/05/2021
							F1905	Device Revision or Replacement	A0203	Defective Device	26/05/2021	26/05/2021
									A070909	Oversensing	26/05/2021	26/05/2021
									A0712	Pacing Problem	26/05/2021	26/05/2021
									A072201	High Impedance	26/05/2021	26/05/2021
11128	2021	All other reportable incidents	36241	Pacemaker/defibrillator lead	CT1684	Cardiac pacemakers and associated devices	E2401	Insufficient Information	A070909	Oversensing		
									A072201	High Impedance		
							F1905	Device Revision or Replacement	A070909	Oversensing		
									A072201	High Impedance		
							F24	Insufficient Health Impact Information	A070909	Oversensing		
									A072201	High Impedance		
11129	2021	All other reportable incidents	35852	Single-chamber implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E2104	Electric Shock	A070909	Oversensing		
									A070910	Undersensing		
									A071302	Inappropriate/Inadequate Shock/Stimulation		
							E2403	No Clinical Signs, Symptoms or Conditions	A070909	Oversensing		
									A070910	Undersensing		
									A071302	Inappropriate/Inadequate Shock/Stimulation		
							F15	Recognized Device or Procedural Complication	A070909	Oversensing		
									A070910	Undersensing		
									A071302	Inappropriate/Inadequate Shock/Stimulation		
11130	2021	All other reportable incidents	35852	Single-chamber implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	21/07/2021	21/07/2021
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	21/07/2021	21/07/2021
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	21/07/2021	21/07/2021
11131	2021	All other reportable incidents	35852	Single-chamber implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	10/08/2021	10/08/2021
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	10/08/2021	10/08/2021
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	10/08/2021	10/08/2021
11132	2021	All other reportable incidents	36241	Pacemaker/defibrillator lead	CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A072201	High Impedance		
							F26	No Health Consequences or Impact	A072201	High Impedance		
11133	2021	All other reportable incidents	35852	Single-chamber implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem	29/06/2021	29/06/2021
									A070504	Premature Discharge of Battery	29/06/2021	29/06/2021
							F08	Hospitalization or Prolonged Hospitalization	A0705	Battery Problem	29/06/2021	29/06/2021
									A070504	Premature Discharge of Battery	29/06/2021	29/06/2021
							F1903	Device Explantation	A0705	Battery Problem	29/06/2021	29/06/2021
									A070504	Premature Discharge of Battery	29/06/2021	29/06/2021
11134	2021	All other reportable incidents	47265	Dual-chamber implantable pacemaker, rate-responsive	CT141	Electrodes/Leads and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery		
							F27	Problem Identified During Non-Clinical Procedure	A070504	Premature Discharge of Battery		
					CT145	Cardiovascular devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery		
							F27	Problem Identified During Non-Clinical Procedure	A070504	Premature Discharge of Battery		
					CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery		
							F27	Problem Identified During Non-Clinical Procedure	A070504	Premature Discharge of Battery		
					CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery		
							F27	Problem Identified During Non-Clinical Procedure	A070504	Premature Discharge of Battery		

					CT233	Surgical	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery		
							F27	Problem Identified During Non-Clinical Procedure	A070504	Premature Discharge of Battery		
					CT2406	Implantable	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery		
							F27	Problem Identified During Non-Clinical Procedure	A070504	Premature Discharge of Battery		
					CT2407	Active implantable	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery		
							F27	Problem Identified During Non-Clinical Procedure	A070504	Premature Discharge of Battery		
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery		
							F27	Problem Identified During Non-Clinical Procedure	A070504	Premature Discharge of Battery		
					CT2727	Active implant devices/ancillaries	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery		
							F27	Problem Identified During Non-Clinical Procedure	A070504	Premature Discharge of Battery		
					CT2750	Implantable cardiac electrical stimulators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery		
							F27	Problem Identified During Non-Clinical Procedure	A070504	Premature Discharge of Battery		
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery		
							F27	Problem Identified During Non-Clinical Procedure	A070504	Premature Discharge of Battery		
					CT2985	Active	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery		
							F27	Problem Identified During Non-Clinical Procedure	A070504	Premature Discharge of Battery		
					CT3001	Implantable metal	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery		
							F27	Problem Identified During Non-Clinical Procedure	A070504	Premature Discharge of Battery		
					CT3002	Implantable synthetic polymer	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery		
							F27	Problem Identified During Non-Clinical Procedure	A070504	Premature Discharge of Battery		
					CT3006	Implantable inorganic material	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery		
							F27	Problem Identified During Non-Clinical Procedure	A070504	Premature Discharge of Battery		
					CT314	Electrically powered	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery		
							F27	Problem Identified During Non-Clinical Procedure	A070504	Premature Discharge of Battery		
					CT321	Long-term surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery		
							F27	Problem Identified During Non-Clinical Procedure	A070504	Premature Discharge of Battery		
					CT449	Generators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery		
							F27	Problem Identified During Non-Clinical Procedure	A070504	Premature Discharge of Battery		
					CT981	Single use	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery		
							F27	Problem Identified During Non-Clinical Procedure	A070504	Premature Discharge of Battery		
					CT983	Surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery		
							F27	Problem Identified During Non-Clinical Procedure	A070504	Premature Discharge of Battery		
11135	2021	All other reportable incidents	35852	Single-chamber implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem	24/06/2021	24/06/2021
									A070504	Premature Discharge of Battery	24/06/2021	24/06/2021
							F1905	Device Revision or Replacement	A0705	Battery Problem	24/06/2021	24/06/2021
									A070504	Premature Discharge of Battery	24/06/2021	24/06/2021
11136	2021	All other reportable incidents	36241	Pacemaker/defibrillator lead	CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070909	Oversensing		
							F13	Misdiagnosis/ Misclassification	A070909	Oversensing		
11137	2021	All other reportable incidents	36241	Pacemaker/defibrillator lead	CT1684	Cardiac pacemakers and associated devices	E2401	Insufficient Information	A070909	Oversensing		
									A072201	High Impedance		
									A090801	Signal Artifact/Noise		
							E2403	No Clinical Signs, Symptoms or Conditions	A070909	Oversensing		
									A072201	High Impedance		
									A090801	Signal Artifact/Noise		
							F15	Recognized Device or Procedural Complication	A070909	Oversensing		
									A072201	High Impedance		
									A090801	Signal Artifact/Noise		
11138	2021	All other reportable incidents	36241	Pacemaker/defibrillator lead	CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A072201	High Impedance		
							F26	No Health Consequences or Impact	A072201	High Impedance		
11139	2021	All other reportable incidents	60910	Coronary venous pacing lead	CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070102	High Capture Threshold		
									A070104	Unstable Capture Threshold		
									A072202	Low Impedance		
							F26	No Health Consequences or Impact	A070102	High Capture Threshold		
									A070104	Unstable Capture Threshold		
									A072202	Low Impedance		
11140	2021	All other reportable incidents	47265	Dual-chamber implantable pacemaker, rate-responsive	CT141	Electrodes/Leads and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	28/05/2021	28/05/2021
							F1901	Additional Surgery	A070504	Premature Discharge of Battery	28/05/2021	28/05/2021
					CT145	Cardiovascular devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	28/05/2021	28/05/2021
							F1901	Additional Surgery	A070504	Premature Discharge of Battery	28/05/2021	28/05/2021
					CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	28/05/2021	28/05/2021
							F1901	Additional Surgery	A070504	Premature Discharge of Battery	28/05/2021	28/05/2021
					CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	28/05/2021	28/05/2021
							F1901	Additional Surgery	A070504	Premature Discharge of Battery	28/05/2021	28/05/2021
					CT233	Surgical	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	28/05/2021	28/05/2021
							F1901	Additional Surgery	A070504	Premature Discharge of Battery	28/05/2021	28/05/2021
					CT2406	Implantable	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	28/05/2021	28/05/2021
							F1901	Additional Surgery	A070504	Premature Discharge of Battery	28/05/2021	28/05/2021
					CT2407	Active implantable	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	28/05/2021	28/05/2021
							F1901	Additional Surgery	A070504	Premature Discharge of Battery	28/05/2021	28/05/2021
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	28/05/2021	28/05/2021
							F1901	Additional Surgery	A070504	Premature Discharge of Battery	28/05/2021	28/05/2021
					CT2727	Active implant devices/ancillaries	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	28/05/2021	28/05/2021

						F1901	Additional Surgery	A070504	Premature Discharge of Battery	28/05/2021	28/05/2021
				CT2750	Implantable cardiac electrical stimulators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	28/05/2021	28/05/2021
						F1901	Additional Surgery	A070504	Premature Discharge of Battery	28/05/2021	28/05/2021
				CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	28/05/2021	28/05/2021
						F1901	Additional Surgery	A070504	Premature Discharge of Battery	28/05/2021	28/05/2021
				CT2985	Active	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	28/05/2021	28/05/2021
						F1901	Additional Surgery	A070504	Premature Discharge of Battery	28/05/2021	28/05/2021
				CT3001	Implantable metal	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	28/05/2021	28/05/2021
						F1901	Additional Surgery	A070504	Premature Discharge of Battery	28/05/2021	28/05/2021
				CT3002	Implantable synthetic polymer	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	28/05/2021	28/05/2021
						F1901	Additional Surgery	A070504	Premature Discharge of Battery	28/05/2021	28/05/2021
				CT3006	Implantable inorganic material	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	28/05/2021	28/05/2021
						F1901	Additional Surgery	A070504	Premature Discharge of Battery	28/05/2021	28/05/2021
				CT314	Electrically powered	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	28/05/2021	28/05/2021
						F1901	Additional Surgery	A070504	Premature Discharge of Battery	28/05/2021	28/05/2021
				CT321	Long-term surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	28/05/2021	28/05/2021
						F1901	Additional Surgery	A070504	Premature Discharge of Battery	28/05/2021	28/05/2021
				CT449	Generators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	28/05/2021	28/05/2021
						F1901	Additional Surgery	A070504	Premature Discharge of Battery	28/05/2021	28/05/2021
				CT981	Single use	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	28/05/2021	28/05/2021
						F1901	Additional Surgery	A070504	Premature Discharge of Battery	28/05/2021	28/05/2021
				CT983	Surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	28/05/2021	28/05/2021
						F1901	Additional Surgery	A070504	Premature Discharge of Battery	28/05/2021	28/05/2021
11141	2021	All other reportable incidents	47270	Cardiac resynchronization therapy implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A0508	Noise, Audible		
									A072201	High Impedance		
							F26	No Health Consequences or Impact	A0508	Noise, Audible		
									A072201	High Impedance		
11142	2021	All other reportable incidents	47265	Dual-chamber implantable pacemaker, rate-responsive	CT141	Electrodes/Leads and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT145	Cardiovascular devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT233	Surgical	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT2406	Implantable	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT2407	Active implantable	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT2727	Active implant devices/ancillaries	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT2750	Implantable cardiac electrical stimulators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT2985	Active	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT3001	Implantable metal	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT3002	Implantable synthetic polymer	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT3006	Implantable inorganic material	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT314	Electrically powered	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT321	Long-term surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT449	Generators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT981	Single use	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT983	Surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
11143	2021	All other reportable incidents	36241	Pacemaker/defibrillator lead	CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A072201	High Impedance		
							F26	No Health Consequences or Impact	A072201	High Impedance		
11144	2021	All other reportable incidents	17882	Physiologic-monitoring defibrillation system	CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070803	Failure to Power Up		
							F27	Problem Identified During Non-Clinical Procedure	A070803	Failure to Power Up		
11145	2021	All other reportable incidents	17882	Physiologic-monitoring defibrillation system	CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070803	Failure to Power Up		
							F27	Problem Identified During Non-Clinical Procedure	A070803	Failure to Power Up		
11146	2021	All other reportable incidents	36241	Pacemaker/defibrillator lead	CT1684	Cardiac pacemakers and associated devices	E2401	Insufficient Information	A072201	High Impedance		
							E2403	No Clinical Signs, Symptoms or Conditions	A072201	High Impedance		
							F26	No Health Consequences or Impact	A072201	High Impedance		
11147	2021	All other reportable incidents	47270	Cardiac resynchronization therapy implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A070909	Oversensing		
									A0712	Pacing Problem		
							F11	Non-Serious Injury/ Illness/ Impairment	A070909	Oversensing		
									A0712	Pacing Problem		
11148	2021	All other reportable incidents	17882	Physiologic-monitoring defibrillation system	CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070803	Failure to Power Up		
							F27	Problem Identified During Non-Clinical Procedure	A070803	Failure to Power Up		
11149	2021	All other reportable incidents	17882	Physiologic-monitoring defibrillation system	CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A1603	Failure of Device to Self-Test		
							F27	Problem Identified During Non-Clinical Procedure	A1603	Failure of Device to Self-Test		
11150	2021	All other reportable incidents	36241	Pacemaker/defibrillator lead	CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070909	Oversensing		
							F26	No Health Consequences or Impact	A070909	Oversensing		
11151	2021	All other reportable incidents	35852	Single-chamber implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A070909	Oversensing		
									A071302	Inappropriate/Inadequate Shock/Stimulation		

							F26	No Health Consequences or Impact	A070909	Oversensing		
									A071302	Inappropriate/Inadequate Shock/Stimulation		
11152	2021	All other reportable incidents	35852	Single-chamber implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A070909	Oversensing		
							F26	No Health Consequences or Impact	A070909	Oversensing		
11153	2021	All other reportable incidents	47270	Cardiac resynchronization therapy implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A070909	Oversensing	25/05/2021	25/05/2021
									A072201	High Impedance	25/05/2021	25/05/2021
									A12	Connection Problem	25/05/2021	25/05/2021
									A1204	Failure to Disconnect	25/05/2021	25/05/2021
									A1205	Loose or Intermittent Connection	25/05/2021	25/05/2021
							F1901	Additional Surgery	A070909	Oversensing	25/05/2021	25/05/2021
									A072201	High Impedance	25/05/2021	25/05/2021
									A12	Connection Problem	25/05/2021	25/05/2021
									A1204	Failure to Disconnect	25/05/2021	25/05/2021
									A1205	Loose or Intermittent Connection	25/05/2021	25/05/2021
							F1905	Device Revision or Replacement	A070909	Oversensing	25/05/2021	25/05/2021
									A072201	High Impedance	25/05/2021	25/05/2021
									A12	Connection Problem	25/05/2021	25/05/2021
									A1204	Failure to Disconnect	25/05/2021	25/05/2021
									A1205	Loose or Intermittent Connection	25/05/2021	25/05/2021
11154	2021	All other reportable incidents	63188	Left-ventricular wireless pacing system sensor/transmitter	CT141	Electrodes/Leads and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070801	Complete Loss of Power		
					CT145	Cardiovascular devices	E2403	No Clinical Signs, Symptoms or Conditions	A070801	Complete Loss of Power		
					CT1577	Wireless monitoring system transmitters	E2403	No Clinical Signs, Symptoms or Conditions	A070801	Complete Loss of Power		
					CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070801	Complete Loss of Power		
					CT2317	Transmitters/Receivers/Transceivers	E2403	No Clinical Signs, Symptoms or Conditions	A070801	Complete Loss of Power		
					CT233	Surgical	E2403	No Clinical Signs, Symptoms or Conditions	A070801	Complete Loss of Power		
					CT2406	Implantable	E2403	No Clinical Signs, Symptoms or Conditions	A070801	Complete Loss of Power		
					CT2407	Active implantable	E2403	No Clinical Signs, Symptoms or Conditions	A070801	Complete Loss of Power		
					CT255	Transducers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070801	Complete Loss of Power		
					CT2726	Physiological monitoring/measuring devices	E2403	No Clinical Signs, Symptoms or Conditions	A070801	Complete Loss of Power		
					CT2727	Active implant devices/ancillaries	E2403	No Clinical Signs, Symptoms or Conditions	A070801	Complete Loss of Power		
					CT2750	Implantable cardiac electrical stimulators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070801	Complete Loss of Power		
					CT2885	Physiological monitoring/measuring system transmitters/receivers/transceivers	E2403	No Clinical Signs, Symptoms or Conditions	A070801	Complete Loss of Power		
					CT2985	Active	E2403	No Clinical Signs, Symptoms or Conditions	A070801	Complete Loss of Power		
					CT3001	Implantable metal	E2403	No Clinical Signs, Symptoms or Conditions	A070801	Complete Loss of Power		
					CT3002	Implantable synthetic polymer	E2403	No Clinical Signs, Symptoms or Conditions	A070801	Complete Loss of Power		
					CT3006	Implantable inorganic material	E2403	No Clinical Signs, Symptoms or Conditions	A070801	Complete Loss of Power		
					CT314	Electrically powered	E2403	No Clinical Signs, Symptoms or Conditions	A070801	Complete Loss of Power		
					CT321	Long-term surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A070801	Complete Loss of Power		
					CT384	Sensors and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070801	Complete Loss of Power		
					CT449	Generators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070801	Complete Loss of Power		
					CT981	Single use	E2403	No Clinical Signs, Symptoms or Conditions	A070801	Complete Loss of Power		
					CT983	Surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A070801	Complete Loss of Power		
					CT994	Sensors	E2403	No Clinical Signs, Symptoms or Conditions	A070801	Complete Loss of Power		
11155	2021	All other reportable incidents	47265	Dual-chamber implantable pacemaker, rate-responsive	CT141	Electrodes/Leads and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem	19/10/2021	19/10/2021
									A070504	Premature Discharge of Battery	19/10/2021	19/10/2021
							F08	Hospitalization or Prolonged Hospitalization	A0705	Battery Problem	19/10/2021	19/10/2021
									A070504	Premature Discharge of Battery	19/10/2021	19/10/2021
							F1905	Device Revision or Replacement	A0705	Battery Problem	19/10/2021	19/10/2021
									A070504	Premature Discharge of Battery	19/10/2021	19/10/2021
					CT145	Cardiovascular devices	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem	19/10/2021	19/10/2021
									A070504	Premature Discharge of Battery	19/10/2021	19/10/2021
							F08	Hospitalization or Prolonged Hospitalization	A0705	Battery Problem	19/10/2021	19/10/2021
									A070504	Premature Discharge of Battery	19/10/2021	19/10/2021
							F1905	Device Revision or Replacement	A0705	Battery Problem	19/10/2021	19/10/2021
									A070504	Premature Discharge of Battery	19/10/2021	19/10/2021
					CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem	19/10/2021	19/10/2021
									A070504	Premature Discharge of Battery	19/10/2021	19/10/2021
							F08	Hospitalization or Prolonged Hospitalization	A0705	Battery Problem	19/10/2021	19/10/2021
									A070504	Premature Discharge of Battery	19/10/2021	19/10/2021
							F1905	Device Revision or Replacement	A0705	Battery Problem	19/10/2021	19/10/2021
									A070504	Premature Discharge of Battery	19/10/2021	19/10/2021
					CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem	19/10/2021	19/10/2021
									A070504	Premature Discharge of Battery	19/10/2021	19/10/2021
							F08	Hospitalization or Prolonged Hospitalization	A0705	Battery Problem	19/10/2021	19/10/2021
									A070504	Premature Discharge of Battery	19/10/2021	19/10/2021
							F1905	Device Revision or Replacement	A0705	Battery Problem	19/10/2021	19/10/2021
									A070504	Premature Discharge of Battery	19/10/2021	19/10/2021
					CT233	Surgical	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem	19/10/2021	19/10/2021
									A070504	Premature Discharge of Battery	19/10/2021	19/10/2021
							F08	Hospitalization or Prolonged Hospitalization	A0705	Battery Problem	19/10/2021	19/10/2021
									A070504	Premature Discharge of Battery	19/10/2021	19/10/2021
							F1905	Device Revision or Replacement	A0705	Battery Problem	19/10/2021	19/10/2021
									A070504	Premature Discharge of Battery	19/10/2021	19/10/2021
					CT2406	Implantable	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem	19/10/2021	19/10/2021
									A070504	Premature Discharge of Battery	19/10/2021	19/10/2021
							F08	Hospitalization or Prolonged Hospitalization	A0705	Battery Problem	19/10/2021	19/10/2021
									A070504	Premature Discharge of Battery	19/10/2021	19/10/2021
							F1905	Device Revision or Replacement	A0705	Battery Problem	19/10/2021	19/10/2021
									A070504	Premature Discharge of Battery	19/10/2021	19/10/2021
					CT2407	Active implantable	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem	19/10/2021	19/10/2021
									A070504	Premature Discharge of Battery	19/10/2021	19/10/2021
							F08	Hospitalization or Prolonged Hospitalization	A0705	Battery Problem	19/10/2021	19/10/2021
									A070504	Premature Discharge of Battery	19/10/2021	19/10/2021
							F1905	Device Revision or Replacement	A0705	Battery Problem	19/10/2021	19/10/2021
									A070504	Premature Discharge of Battery	19/10/2021	19/10/2021
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem	19/10/2021	19/10/2021

									A070504	Premature Discharge of Battery	19/10/2021	19/10/2021
							F08	Hospitalization or Prolonged Hospitalization	A0705	Battery Problem	19/10/2021	19/10/2021
									A070504	Premature Discharge of Battery	19/10/2021	19/10/2021
							F1905	Device Revision or Replacement	A0705	Battery Problem	19/10/2021	19/10/2021
									A070504	Premature Discharge of Battery	19/10/2021	19/10/2021
					CT2727	Active implant devices/ancillaries	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem	19/10/2021	19/10/2021
									A070504	Premature Discharge of Battery	19/10/2021	19/10/2021
							F08	Hospitalization or Prolonged Hospitalization	A0705	Battery Problem	19/10/2021	19/10/2021
									A070504	Premature Discharge of Battery	19/10/2021	19/10/2021
							F1905	Device Revision or Replacement	A0705	Battery Problem	19/10/2021	19/10/2021
									A070504	Premature Discharge of Battery	19/10/2021	19/10/2021
					CT2750	Implantable cardiac electrical stimulators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem	19/10/2021	19/10/2021
									A070504	Premature Discharge of Battery	19/10/2021	19/10/2021
							F08	Hospitalization or Prolonged Hospitalization	A0705	Battery Problem	19/10/2021	19/10/2021
									A070504	Premature Discharge of Battery	19/10/2021	19/10/2021
							F1905	Device Revision or Replacement	A0705	Battery Problem	19/10/2021	19/10/2021
									A070504	Premature Discharge of Battery	19/10/2021	19/10/2021
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem	19/10/2021	19/10/2021
									A070504	Premature Discharge of Battery	19/10/2021	19/10/2021
							F08	Hospitalization or Prolonged Hospitalization	A0705	Battery Problem	19/10/2021	19/10/2021
									A070504	Premature Discharge of Battery	19/10/2021	19/10/2021
							F1905	Device Revision or Replacement	A0705	Battery Problem	19/10/2021	19/10/2021
									A070504	Premature Discharge of Battery	19/10/2021	19/10/2021
					CT2985	Active	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem	19/10/2021	19/10/2021
									A070504	Premature Discharge of Battery	19/10/2021	19/10/2021
							F08	Hospitalization or Prolonged Hospitalization	A0705	Battery Problem	19/10/2021	19/10/2021
									A070504	Premature Discharge of Battery	19/10/2021	19/10/2021
							F1905	Device Revision or Replacement	A0705	Battery Problem	19/10/2021	19/10/2021
									A070504	Premature Discharge of Battery	19/10/2021	19/10/2021
					CT3001	Implantable metal	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem	19/10/2021	19/10/2021
									A070504	Premature Discharge of Battery	19/10/2021	19/10/2021
							F08	Hospitalization or Prolonged Hospitalization	A0705	Battery Problem	19/10/2021	19/10/2021
									A070504	Premature Discharge of Battery	19/10/2021	19/10/2021
							F1905	Device Revision or Replacement	A0705	Battery Problem	19/10/2021	19/10/2021
									A070504	Premature Discharge of Battery	19/10/2021	19/10/2021
					CT3002	Implantable synthetic polymer	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem	19/10/2021	19/10/2021
									A070504	Premature Discharge of Battery	19/10/2021	19/10/2021
							F08	Hospitalization or Prolonged Hospitalization	A0705	Battery Problem	19/10/2021	19/10/2021
									A070504	Premature Discharge of Battery	19/10/2021	19/10/2021
							F1905	Device Revision or Replacement	A0705	Battery Problem	19/10/2021	19/10/2021
									A070504	Premature Discharge of Battery	19/10/2021	19/10/2021
					CT3006	Implantable inorganic material	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem	19/10/2021	19/10/2021
									A070504	Premature Discharge of Battery	19/10/2021	19/10/2021
							F08	Hospitalization or Prolonged Hospitalization	A0705	Battery Problem	19/10/2021	19/10/2021
									A070504	Premature Discharge of Battery	19/10/2021	19/10/2021
							F1905	Device Revision or Replacement	A0705	Battery Problem	19/10/2021	19/10/2021
									A070504	Premature Discharge of Battery	19/10/2021	19/10/2021
					CT314	Electrically powered	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem	19/10/2021	19/10/2021
									A070504	Premature Discharge of Battery	19/10/2021	19/10/2021
							F08	Hospitalization or Prolonged Hospitalization	A0705	Battery Problem	19/10/2021	19/10/2021
									A070504	Premature Discharge of Battery	19/10/2021	19/10/2021
							F1905	Device Revision or Replacement	A0705	Battery Problem	19/10/2021	19/10/2021
									A070504	Premature Discharge of Battery	19/10/2021	19/10/2021
					CT321	Long-term surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem	19/10/2021	19/10/2021
									A070504	Premature Discharge of Battery	19/10/2021	19/10/2021
							F08	Hospitalization or Prolonged Hospitalization	A0705	Battery Problem	19/10/2021	19/10/2021
									A070504	Premature Discharge of Battery	19/10/2021	19/10/2021
							F1905	Device Revision or Replacement	A0705	Battery Problem	19/10/2021	19/10/2021
									A070504	Premature Discharge of Battery	19/10/2021	19/10/2021
					CT449	Generators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem	19/10/2021	19/10/2021
									A070504	Premature Discharge of Battery	19/10/2021	19/10/2021
							F08	Hospitalization or Prolonged Hospitalization	A0705	Battery Problem	19/10/2021	19/10/2021
									A070504	Premature Discharge of Battery	19/10/2021	19/10/2021
							F1905	Device Revision or Replacement	A0705	Battery Problem	19/10/2021	19/10/2021
									A070504	Premature Discharge of Battery	19/10/2021	19/10/2021
					CT981	Single use	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem	19/10/2021	19/10/2021
									A070504	Premature Discharge of Battery	19/10/2021	19/10/2021
							F08	Hospitalization or Prolonged Hospitalization	A0705	Battery Problem	19/10/2021	19/10/2021
									A070504	Premature Discharge of Battery	19/10/2021	19/10/2021
							F1905	Device Revision or Replacement	A0705	Battery Problem	19/10/2021	19/10/2021
									A070504	Premature Discharge of Battery	19/10/2021	19/10/2021
					CT983	Surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem	19/10/2021	19/10/2021
									A070504	Premature Discharge of Battery	19/10/2021	19/10/2021
							F08	Hospitalization or Prolonged Hospitalization	A0705	Battery Problem	19/10/2021	19/10/2021
									A070504	Premature Discharge of Battery	19/10/2021	19/10/2021
							F1905	Device Revision or Replacement	A0705	Battery Problem	19/10/2021	19/10/2021
									A070504	Premature Discharge of Battery	19/10/2021	19/10/2021
11156	2021	All other reportable incidents	36241	Pacemaker/defibrillator lead	CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070909	Oversensing		
									A0712	Pacing Problem		
							F11	Non-Serious Injury/ Illness/ Impairment	A070909	Oversensing		
									A0712	Pacing Problem		
11157	2021	All other reportable incidents	36241	Pacemaker/defibrillator lead	CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070909	Oversensing		
									A0712	Pacing Problem		
							F11	Non-Serious Injury/ Illness/ Impairment	A070909	Oversensing		
									A0712	Pacing Problem		
11158	2021	All other reportable incidents	37265	Dual-chamber implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem	23/06/2021	23/06/2021
									A070504	Premature Discharge of Battery	23/06/2021	23/06/2021
							F08	Hospitalization or Prolonged Hospitalization	A0705	Battery Problem	23/06/2021	23/06/2021

								A070504	Premature Discharge of Battery	23/06/2021	23/06/2021
						F1905	Device Revision or Replacement	A0705	Battery Problem	23/06/2021	23/06/2021
								A070504	Premature Discharge of Battery	23/06/2021	23/06/2021
11159	2021	All other reportable incidents	47265	Dual-chamber implantable pacemaker, rate-responsive	CT141	Electrodes/Leads and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT145	Cardiovascular devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT233	Surgical	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT2406	Implantable	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT2407	Active implantable	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT2727	Active implant devices/ancillaries	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT2750	Implantable cardiac electrical stimulators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT2985	Active	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT3001	Implantable metal	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT3002	Implantable synthetic polymer	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT3006	Implantable inorganic material	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT314	Electrically powered	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT321	Long-term surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT449	Generators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT981	Single use	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT983	Surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
11160	2021	All other reportable incidents	47265	Dual-chamber implantable pacemaker, rate-responsive	CT141	Electrodes/Leads and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT145	Cardiovascular devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT233	Surgical	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT2406	Implantable	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT2407	Active implantable	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT2727	Active implant devices/ancillaries	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT2750	Implantable cardiac electrical stimulators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT2985	Active	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT3001	Implantable metal	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT3002	Implantable synthetic polymer	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT3006	Implantable inorganic material	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT314	Electrically powered	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT321	Long-term surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT449	Generators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT981	Single use	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT983	Surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
11161	2021	All other reportable incidents	47265	Dual-chamber implantable pacemaker, rate-responsive	CT141	Electrodes/Leads and associated devices	E060110	Ventricular Fibrillation	A071202	Inaccurate Synchronization		
							E2403	No Clinical Signs, Symptoms or Conditions	A071202	Inaccurate Synchronization		
							F12	Serious Injury/ Illness/ Impairment	A071202	Inaccurate Synchronization		
					CT145	Cardiovascular devices	E060110	Ventricular Fibrillation	A071202	Inaccurate Synchronization		
							E2403	No Clinical Signs, Symptoms or Conditions	A071202	Inaccurate Synchronization		

							F12	Serious Injury/ Illness/ Impairment	A071202	Inaccurate Synchronization		
					CT1684	Cardiac pacemakers and associated devices	E060110	Ventricular Fibrillation	A071202	Inaccurate Synchronization		
							E2403	No Clinical Signs, Symptoms or Conditions	A071202	Inaccurate Synchronization		
							F12	Serious Injury/ Illness/ Impairment	A071202	Inaccurate Synchronization		
					CT1685	Cardiac pacemakers	E060110	Ventricular Fibrillation	A071202	Inaccurate Synchronization		
							E2403	No Clinical Signs, Symptoms or Conditions	A071202	Inaccurate Synchronization		
							F12	Serious Injury/ Illness/ Impairment	A071202	Inaccurate Synchronization		
					CT233	Surgical	E060110	Ventricular Fibrillation	A071202	Inaccurate Synchronization		
							E2403	No Clinical Signs, Symptoms or Conditions	A071202	Inaccurate Synchronization		
							F12	Serious Injury/ Illness/ Impairment	A071202	Inaccurate Synchronization		
					CT2406	Implantable	E060110	Ventricular Fibrillation	A071202	Inaccurate Synchronization		
							E2403	No Clinical Signs, Symptoms or Conditions	A071202	Inaccurate Synchronization		
							F12	Serious Injury/ Illness/ Impairment	A071202	Inaccurate Synchronization		
					CT2407	Active implantable	E060110	Ventricular Fibrillation	A071202	Inaccurate Synchronization		
							E2403	No Clinical Signs, Symptoms or Conditions	A071202	Inaccurate Synchronization		
							F12	Serious Injury/ Illness/ Impairment	A071202	Inaccurate Synchronization		
					CT257	Implantable cardiac pacemakers	E060110	Ventricular Fibrillation	A071202	Inaccurate Synchronization		
							E2403	No Clinical Signs, Symptoms or Conditions	A071202	Inaccurate Synchronization		
							F12	Serious Injury/ Illness/ Impairment	A071202	Inaccurate Synchronization		
					CT2727	Active implant devices/ancillaries	E060110	Ventricular Fibrillation	A071202	Inaccurate Synchronization		
							E2403	No Clinical Signs, Symptoms or Conditions	A071202	Inaccurate Synchronization		
							F12	Serious Injury/ Illness/ Impairment	A071202	Inaccurate Synchronization		
					CT2750	Implantable cardiac electrical stimulators and associated devices	E060110	Ventricular Fibrillation	A071202	Inaccurate Synchronization		
							E2403	No Clinical Signs, Symptoms or Conditions	A071202	Inaccurate Synchronization		
							F12	Serious Injury/ Illness/ Impairment	A071202	Inaccurate Synchronization		
					CT2866	Implantable cardiac electrical stimulators	E060110	Ventricular Fibrillation	A071202	Inaccurate Synchronization		
							E2403	No Clinical Signs, Symptoms or Conditions	A071202	Inaccurate Synchronization		
							F12	Serious Injury/ Illness/ Impairment	A071202	Inaccurate Synchronization		
					CT2985	Active	E060110	Ventricular Fibrillation	A071202	Inaccurate Synchronization		
							E2403	No Clinical Signs, Symptoms or Conditions	A071202	Inaccurate Synchronization		
							F12	Serious Injury/ Illness/ Impairment	A071202	Inaccurate Synchronization		
					CT3001	Implantable metal	E060110	Ventricular Fibrillation	A071202	Inaccurate Synchronization		
							E2403	No Clinical Signs, Symptoms or Conditions	A071202	Inaccurate Synchronization		
							F12	Serious Injury/ Illness/ Impairment	A071202	Inaccurate Synchronization		
					CT3002	Implantable synthetic polymer	E060110	Ventricular Fibrillation	A071202	Inaccurate Synchronization		
							E2403	No Clinical Signs, Symptoms or Conditions	A071202	Inaccurate Synchronization		
							F12	Serious Injury/ Illness/ Impairment	A071202	Inaccurate Synchronization		
					CT3006	Implantable inorganic material	E060110	Ventricular Fibrillation	A071202	Inaccurate Synchronization		
							E2403	No Clinical Signs, Symptoms or Conditions	A071202	Inaccurate Synchronization		
							F12	Serious Injury/ Illness/ Impairment	A071202	Inaccurate Synchronization		
					CT314	Electrically powered	E060110	Ventricular Fibrillation	A071202	Inaccurate Synchronization		
							E2403	No Clinical Signs, Symptoms or Conditions	A071202	Inaccurate Synchronization		
							F12	Serious Injury/ Illness/ Impairment	A071202	Inaccurate Synchronization		
					CT321	Long-term surgical invasive	E060110	Ventricular Fibrillation	A071202	Inaccurate Synchronization		
							E2403	No Clinical Signs, Symptoms or Conditions	A071202	Inaccurate Synchronization		
							F12	Serious Injury/ Illness/ Impairment	A071202	Inaccurate Synchronization		
					CT449	Generators and associated devices	E060110	Ventricular Fibrillation	A071202	Inaccurate Synchronization		
							E2403	No Clinical Signs, Symptoms or Conditions	A071202	Inaccurate Synchronization		
							F12	Serious Injury/ Illness/ Impairment	A071202	Inaccurate Synchronization		
					CT981	Single use	E060110	Ventricular Fibrillation	A071202	Inaccurate Synchronization		
							E2403	No Clinical Signs, Symptoms or Conditions	A071202	Inaccurate Synchronization		
							F12	Serious Injury/ Illness/ Impairment	A071202	Inaccurate Synchronization		
					CT983	Surgical invasive	E060110	Ventricular Fibrillation	A071202	Inaccurate Synchronization		
							E2403	No Clinical Signs, Symptoms or Conditions	A071202	Inaccurate Synchronization		
							F12	Serious Injury/ Illness/ Impairment	A071202	Inaccurate Synchronization		
11162	2021	All other reportable incidents	35852	Single-chamber implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem	07/10/2021	07/10/2021
									A070504	Premature Discharge of Battery	07/10/2021	07/10/2021
							F08	Hospitalization or Prolonged Hospitalization	A0705	Battery Problem	07/10/2021	07/10/2021
									A070504	Premature Discharge of Battery	07/10/2021	07/10/2021
							F1905	Device Revision or Replacement	A0705	Battery Problem	07/10/2021	07/10/2021
									A070504	Premature Discharge of Battery	07/10/2021	07/10/2021
11163	2021	All other reportable incidents	17882	Physiologic-monitoring defibrillation system	CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070908	Failure to Sense		
							F26	No Health Consequences or Impact	A070908	Failure to Sense		
11164	2021	All other reportable incidents	35852	Single-chamber implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions				
11165	2021	All other reportable incidents	35223	Endocardial/interventricular septal pacing lead	CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A0401	Break		
									A070909	Oversensing		
							F1901	Additional Surgery	A0401	Break		
									A070909	Oversensing		
							F1905	Device Revision or Replacement	A0401	Break		
									A070909	Oversensing		
11166	2021	All other reportable incidents	47270	Cardiac resynchronization therapy implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A070101	Failure to Capture		
									A072201	High Impedance		
									A12	Connection Problem		
							F08	Hospitalization or Prolonged Hospitalization	A070101	Failure to Capture		
									A072201	High Impedance		
									A12	Connection Problem		
							F12	Serious Injury/ Illness/ Impairment	A070101	Failure to Capture		
									A072201	High Impedance		
									A12	Connection Problem		
							F1905	Device Revision or Replacement	A070101	Failure to Capture		
									A072201	High Impedance		
									A12	Connection Problem		
11167	2021	All other reportable incidents	36241	Pacemaker/defibrillator lead	CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070101	Failure to Capture		
							F08	Hospitalization or Prolonged Hospitalization	A070101	Failure to Capture		
							F12	Serious Injury/ Illness/ Impairment	A070101	Failure to Capture		
							F1905	Device Revision or Replacement	A070101	Failure to Capture		
11168	2021	All other reportable incidents	37265	Dual-chamber implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E2308	Deformity/ Disfigurement	A0722	Impedance Problem		
							E2311	Discomfort	A0722	Impedance Problem		

							F11	Non-Serious Injury/ Illness/ Impairment	A0722	Impedance Problem		
11169	2021	All other reportable incidents	47267	Single-chamber implantable pacemaker, rate-responsive	CT141	Electrodes/Leads and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	28/05/2021	28/05/2021
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	28/05/2021	28/05/2021
							F1903	Device Explantation	A070504	Premature Discharge of Battery	28/05/2021	28/05/2021
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	28/05/2021	28/05/2021
					CT145	Cardiovascular devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	28/05/2021	28/05/2021
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	28/05/2021	28/05/2021
							F1903	Device Explantation	A070504	Premature Discharge of Battery	28/05/2021	28/05/2021
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	28/05/2021	28/05/2021
					CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	28/05/2021	28/05/2021
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	28/05/2021	28/05/2021
							F1903	Device Explantation	A070504	Premature Discharge of Battery	28/05/2021	28/05/2021
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	28/05/2021	28/05/2021
					CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	28/05/2021	28/05/2021
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	28/05/2021	28/05/2021
							F1903	Device Explantation	A070504	Premature Discharge of Battery	28/05/2021	28/05/2021
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	28/05/2021	28/05/2021
					CT233	Surgical	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	28/05/2021	28/05/2021
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	28/05/2021	28/05/2021
							F1903	Device Explantation	A070504	Premature Discharge of Battery	28/05/2021	28/05/2021
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	28/05/2021	28/05/2021
					CT2406	Implantable	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	28/05/2021	28/05/2021
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	28/05/2021	28/05/2021
							F1903	Device Explantation	A070504	Premature Discharge of Battery	28/05/2021	28/05/2021
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	28/05/2021	28/05/2021
					CT2407	Active implantable	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	28/05/2021	28/05/2021
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	28/05/2021	28/05/2021
							F1903	Device Explantation	A070504	Premature Discharge of Battery	28/05/2021	28/05/2021
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	28/05/2021	28/05/2021
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	28/05/2021	28/05/2021
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	28/05/2021	28/05/2021
							F1903	Device Explantation	A070504	Premature Discharge of Battery	28/05/2021	28/05/2021
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	28/05/2021	28/05/2021
					CT2727	Active implant devices/ancillaries	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	28/05/2021	28/05/2021
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	28/05/2021	28/05/2021
							F1903	Device Explantation	A070504	Premature Discharge of Battery	28/05/2021	28/05/2021
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	28/05/2021	28/05/2021
					CT2750	Implantable cardiac electrical stimulators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	28/05/2021	28/05/2021
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	28/05/2021	28/05/2021
							F1903	Device Explantation	A070504	Premature Discharge of Battery	28/05/2021	28/05/2021
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	28/05/2021	28/05/2021
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	28/05/2021	28/05/2021
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	28/05/2021	28/05/2021
							F1903	Device Explantation	A070504	Premature Discharge of Battery	28/05/2021	28/05/2021
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	28/05/2021	28/05/2021
					CT2985	Active	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	28/05/2021	28/05/2021
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	28/05/2021	28/05/2021
							F1903	Device Explantation	A070504	Premature Discharge of Battery	28/05/2021	28/05/2021
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	28/05/2021	28/05/2021
					CT3001	Implantable metal	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	28/05/2021	28/05/2021
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	28/05/2021	28/05/2021
							F1903	Device Explantation	A070504	Premature Discharge of Battery	28/05/2021	28/05/2021
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	28/05/2021	28/05/2021
					CT3002	Implantable synthetic polymer	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	28/05/2021	28/05/2021
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	28/05/2021	28/05/2021
							F1903	Device Explantation	A070504	Premature Discharge of Battery	28/05/2021	28/05/2021
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	28/05/2021	28/05/2021
					CT3006	Implantable inorganic material	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	28/05/2021	28/05/2021
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	28/05/2021	28/05/2021
							F1903	Device Explantation	A070504	Premature Discharge of Battery	28/05/2021	28/05/2021
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	28/05/2021	28/05/2021
					CT314	Electrically powered	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	28/05/2021	28/05/2021
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	28/05/2021	28/05/2021
							F1903	Device Explantation	A070504	Premature Discharge of Battery	28/05/2021	28/05/2021
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	28/05/2021	28/05/2021
					CT321	Long-term surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	28/05/2021	28/05/2021
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	28/05/2021	28/05/2021
							F1903	Device Explantation	A070504	Premature Discharge of Battery	28/05/2021	28/05/2021
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	28/05/2021	28/05/2021
					CT449	Generators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	28/05/2021	28/05/2021
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	28/05/2021	28/05/2021
							F1903	Device Explantation	A070504	Premature Discharge of Battery	28/05/2021	28/05/2021
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	28/05/2021	28/05/2021
					CT981	Single use	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	28/05/2021	28/05/2021
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	28/05/2021	28/05/2021
							F1903	Device Explantation	A070504	Premature Discharge of Battery	28/05/2021	28/05/2021
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	28/05/2021	28/05/2021
					CT983	Surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	28/05/2021	28/05/2021
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	28/05/2021	28/05/2021
							F1903	Device Explantation	A070504	Premature Discharge of Battery	28/05/2021	28/05/2021
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	28/05/2021	28/05/2021
11170	2021	All other reportable incidents	17882	Physiologic-monitoring defibrillation system	CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A1603	Failure of Device to Self-Test		
							F27	Problem Identified During Non-Clinical Procedure	A1603	Failure of Device to Self-Test		
11171	2021	All other reportable incidents	37265	Dual-chamber implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	04/06/2021	04/06/2021
							F1901	Additional Surgery	A070504	Premature Discharge of Battery	04/06/2021	04/06/2021
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	04/06/2021	04/06/2021
11172	2021	All other reportable incidents	36241	Pacemaker/defibrillator lead	CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070909	Oversensing		
									A071302	Inappropriate/Inadequate Shock/Stimulation		

						F10	Inadequate/inappropriate Treatment or Diagnostic Exposure	A070909	Oversensing			
								A071302	Inappropriate/Inadequate Shock/Stimulation			
						F19	Surgical Intervention	A070909	Oversensing			
								A071302	Inappropriate/Inadequate Shock/Stimulation			
11173	2021	All other reportable incidents	35852	Single-chamber implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A070102	High Capture Threshold		
									A072201	High Impedance		
									A1601	Device Alarm System		
							F26	No Health Consequences or Impact	A070102	High Capture Threshold		
									A072201	High Impedance		
									A1601	Device Alarm System		
11174	2021	All other reportable incidents	35852	Single-chamber implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	14/07/2021	14/07/2021
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	14/07/2021	14/07/2021
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	14/07/2021	14/07/2021
11175	2021	All other reportable incidents	47265	Dual-chamber implantable pacemaker, rate-responsive	CT141	Electrodes/Leads and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	18/06/2021	18/06/2021
							F1901	Additional Surgery	A070504	Premature Discharge of Battery	18/06/2021	18/06/2021
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	18/06/2021	18/06/2021
					CT145	Cardiovascular devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	18/06/2021	18/06/2021
							F1901	Additional Surgery	A070504	Premature Discharge of Battery	18/06/2021	18/06/2021
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	18/06/2021	18/06/2021
					CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	18/06/2021	18/06/2021
							F1901	Additional Surgery	A070504	Premature Discharge of Battery	18/06/2021	18/06/2021
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	18/06/2021	18/06/2021
					CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	18/06/2021	18/06/2021
							F1901	Additional Surgery	A070504	Premature Discharge of Battery	18/06/2021	18/06/2021
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	18/06/2021	18/06/2021
					CT233	Surgical	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	18/06/2021	18/06/2021
							F1901	Additional Surgery	A070504	Premature Discharge of Battery	18/06/2021	18/06/2021
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	18/06/2021	18/06/2021
					CT2406	Implantable	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	18/06/2021	18/06/2021
							F1901	Additional Surgery	A070504	Premature Discharge of Battery	18/06/2021	18/06/2021
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	18/06/2021	18/06/2021
					CT2407	Active implantable	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	18/06/2021	18/06/2021
							F1901	Additional Surgery	A070504	Premature Discharge of Battery	18/06/2021	18/06/2021
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	18/06/2021	18/06/2021
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	18/06/2021	18/06/2021
							F1901	Additional Surgery	A070504	Premature Discharge of Battery	18/06/2021	18/06/2021
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	18/06/2021	18/06/2021
					CT2727	Active implant devices/ancillaries	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	18/06/2021	18/06/2021
							F1901	Additional Surgery	A070504	Premature Discharge of Battery	18/06/2021	18/06/2021
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	18/06/2021	18/06/2021
					CT2750	Implantable cardiac electrical stimulators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	18/06/2021	18/06/2021
							F1901	Additional Surgery	A070504	Premature Discharge of Battery	18/06/2021	18/06/2021
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	18/06/2021	18/06/2021
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	18/06/2021	18/06/2021
							F1901	Additional Surgery	A070504	Premature Discharge of Battery	18/06/2021	18/06/2021
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	18/06/2021	18/06/2021
					CT2985	Active	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	18/06/2021	18/06/2021
							F1901	Additional Surgery	A070504	Premature Discharge of Battery	18/06/2021	18/06/2021
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	18/06/2021	18/06/2021
					CT3001	Implantable metal	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	18/06/2021	18/06/2021
							F1901	Additional Surgery	A070504	Premature Discharge of Battery	18/06/2021	18/06/2021
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	18/06/2021	18/06/2021
					CT3002	Implantable synthetic polymer	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	18/06/2021	18/06/2021
							F1901	Additional Surgery	A070504	Premature Discharge of Battery	18/06/2021	18/06/2021
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	18/06/2021	18/06/2021
					CT3006	Implantable inorganic material	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	18/06/2021	18/06/2021
							F1901	Additional Surgery	A070504	Premature Discharge of Battery	18/06/2021	18/06/2021
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	18/06/2021	18/06/2021
					CT314	Electrically powered	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	18/06/2021	18/06/2021
							F1901	Additional Surgery	A070504	Premature Discharge of Battery	18/06/2021	18/06/2021
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	18/06/2021	18/06/2021
					CT321	Long-term surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	18/06/2021	18/06/2021
							F1901	Additional Surgery	A070504	Premature Discharge of Battery	18/06/2021	18/06/2021
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	18/06/2021	18/06/2021
					CT449	Generators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	18/06/2021	18/06/2021
							F1901	Additional Surgery	A070504	Premature Discharge of Battery	18/06/2021	18/06/2021
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	18/06/2021	18/06/2021
					CT981	Single use	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	18/06/2021	18/06/2021
							F1901	Additional Surgery	A070504	Premature Discharge of Battery	18/06/2021	18/06/2021
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	18/06/2021	18/06/2021
					CT983	Surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	18/06/2021	18/06/2021
							F1901	Additional Surgery	A070504	Premature Discharge of Battery	18/06/2021	18/06/2021
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	18/06/2021	18/06/2021
11176	2021	All other reportable incidents	35852	Single-chamber implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E2007	Fall	A010402	Migration	26/05/2021	26/05/2021
									A0203	Defective Device	26/05/2021	26/05/2021
							F08	Hospitalization or Prolonged Hospitalization	A010402	Migration	26/05/2021	26/05/2021
									A0203	Defective Device	26/05/2021	26/05/2021
							F12	Serious Injury/ Illness/ Impairment	A010402	Migration	26/05/2021	26/05/2021
									A0203	Defective Device	26/05/2021	26/05/2021
							F19	Surgical Intervention	A010402	Migration	26/05/2021	26/05/2021
									A0203	Defective Device	26/05/2021	26/05/2021
							F1905	Device Revision or Replacement	A010402	Migration	26/05/2021	26/05/2021
									A0203	Defective Device	26/05/2021	26/05/2021
11177	2021	All other reportable incidents	17882	Physiologic-monitoring defibrillation system	CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070908	Failure to Sense		
							F27	Problem Identified During Non-Clinical Procedure	A070908	Failure to Sense		
11178	2021	All other reportable incidents	36241	Pacemaker/defibrillator lead	CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A0203	Defective Device	26/05/2021	26/05/2021
									A051201	Device Dislodged or Dislocated	26/05/2021	26/05/2021
							F08	Hospitalization or Prolonged Hospitalization	A0203	Defective Device	26/05/2021	26/05/2021

									A051201	Device Dislodged or Dislocated	26/05/2021	26/05/2021
							F19	Surgical Intervention	A0203	Defective Device	26/05/2021	26/05/2021
									A051201	Device Dislodged or Dislocated	26/05/2021	26/05/2021
11179	2021	All other reportable incidents	35852	Single-chamber implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	12/07/2021	12/07/2021
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	12/07/2021	12/07/2021
							F1903	Device Explantation	A070504	Premature Discharge of Battery	12/07/2021	12/07/2021
11180	2021	All other reportable incidents	37265	Dual-chamber implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem	10/08/2021	10/08/2021
									A070504	Premature Discharge of Battery	10/08/2021	10/08/2021
							F26	No Health Consequences or Impact	A0705	Battery Problem	10/08/2021	10/08/2021
									A070504	Premature Discharge of Battery	10/08/2021	10/08/2021
11181	2021	All other reportable incidents	35852	Single-chamber implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	28/06/2021	28/06/2021
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery	28/06/2021	28/06/2021
11182	2021	All other reportable incidents	47270	Cardiac resynchronization therapy implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	03/06/2021	03/06/2021
							F02	Death	A070504	Premature Discharge of Battery	03/06/2021	03/06/2021
11183	2021	All other reportable incidents	47265	Dual-chamber implantable pacemaker, rate-responsive	CT141	Electrodes/Leads and associated devices	E0112	Dizziness	A070909	Oversensing		
									A071204	No Pacing		
							F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A070909	Oversensing		
									A071204	No Pacing		
							F11	Non-Serious Injury/ Illness/ Impairment	A070909	Oversensing		
									A071204	No Pacing		
					CT145	Cardiovascular devices	E0112	Dizziness	A070909	Oversensing		
									A071204	No Pacing		
							F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A070909	Oversensing		
									A071204	No Pacing		
							F11	Non-Serious Injury/ Illness/ Impairment	A070909	Oversensing		
									A071204	No Pacing		
					CT1684	Cardiac pacemakers and associated devices	E0112	Dizziness	A070909	Oversensing		
									A071204	No Pacing		
							F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A070909	Oversensing		
									A071204	No Pacing		
							F11	Non-Serious Injury/ Illness/ Impairment	A070909	Oversensing		
									A071204	No Pacing		
					CT1685	Cardiac pacemakers	E0112	Dizziness	A070909	Oversensing		
									A071204	No Pacing		
							F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A070909	Oversensing		
									A071204	No Pacing		
							F11	Non-Serious Injury/ Illness/ Impairment	A070909	Oversensing		
									A071204	No Pacing		
					CT233	Surgical	E0112	Dizziness	A070909	Oversensing		
									A071204	No Pacing		
							F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A070909	Oversensing		
									A071204	No Pacing		
							F11	Non-Serious Injury/ Illness/ Impairment	A070909	Oversensing		
									A071204	No Pacing		
					CT2406	Implantable	E0112	Dizziness	A070909	Oversensing		
									A071204	No Pacing		
							F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A070909	Oversensing		
									A071204	No Pacing		
							F11	Non-Serious Injury/ Illness/ Impairment	A070909	Oversensing		
									A071204	No Pacing		
					CT2407	Active implantable	E0112	Dizziness	A070909	Oversensing		
									A071204	No Pacing		
							F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A070909	Oversensing		
									A071204	No Pacing		
							F11	Non-Serious Injury/ Illness/ Impairment	A070909	Oversensing		
									A071204	No Pacing		
					CT257	Implantable cardiac pacemakers	E0112	Dizziness	A070909	Oversensing		
									A071204	No Pacing		
							F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A070909	Oversensing		
									A071204	No Pacing		
							F11	Non-Serious Injury/ Illness/ Impairment	A070909	Oversensing		
									A071204	No Pacing		
					CT2727	Active implant devices/ancillaries	E0112	Dizziness	A070909	Oversensing		
									A071204	No Pacing		
							F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A070909	Oversensing		
									A071204	No Pacing		
							F11	Non-Serious Injury/ Illness/ Impairment	A070909	Oversensing		
									A071204	No Pacing		
					CT2750	Implantable cardiac electrical stimulators and associated devices	E0112	Dizziness	A070909	Oversensing		
									A071204	No Pacing		
							F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A070909	Oversensing		
									A071204	No Pacing		
							F11	Non-Serious Injury/ Illness/ Impairment	A070909	Oversensing		
									A071204	No Pacing		
					CT2866	Implantable cardiac electrical stimulators	E0112	Dizziness	A070909	Oversensing		
									A071204	No Pacing		
							F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A070909	Oversensing		
									A071204	No Pacing		
							F11	Non-Serious Injury/ Illness/ Impairment	A070909	Oversensing		
									A071204	No Pacing		
					CT2985	Active	E0112	Dizziness	A070909	Oversensing		
									A071204	No Pacing		
							F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A070909	Oversensing		
									A071204	No Pacing		
							F11	Non-Serious Injury/ Illness/ Impairment	A070909	Oversensing		
									A071204	No Pacing		
					CT3001	Implantable metal	E0112	Dizziness	A070909	Oversensing		
									A071204	No Pacing		

							F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A070909	Oversensing		
									A071204	No Pacing		
							F11	Non-Serious Injury/ Illness/ Impairment	A070909	Oversensing		
									A071204	No Pacing		
					CT3002	Implantable synthetic polymer	E0112	Dizziness	A070909	Oversensing		
									A071204	No Pacing		
							F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A070909	Oversensing		
									A071204	No Pacing		
							F11	Non-Serious Injury/ Illness/ Impairment	A070909	Oversensing		
									A071204	No Pacing		
					CT3006	Implantable inorganic material	E0112	Dizziness	A070909	Oversensing		
									A071204	No Pacing		
							F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A070909	Oversensing		
									A071204	No Pacing		
							F11	Non-Serious Injury/ Illness/ Impairment	A070909	Oversensing		
									A071204	No Pacing		
					CT314	Electrically powered	E0112	Dizziness	A070909	Oversensing		
									A071204	No Pacing		
							F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A070909	Oversensing		
									A071204	No Pacing		
							F11	Non-Serious Injury/ Illness/ Impairment	A070909	Oversensing		
									A071204	No Pacing		
					CT321	Long-term surgical invasive	E0112	Dizziness	A070909	Oversensing		
									A071204	No Pacing		
							F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A070909	Oversensing		
									A071204	No Pacing		
							F11	Non-Serious Injury/ Illness/ Impairment	A070909	Oversensing		
									A071204	No Pacing		
					CT449	Generators and associated devices	E0112	Dizziness	A070909	Oversensing		
									A071204	No Pacing		
							F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A070909	Oversensing		
									A071204	No Pacing		
							F11	Non-Serious Injury/ Illness/ Impairment	A070909	Oversensing		
									A071204	No Pacing		
					CT981	Single use	E0112	Dizziness	A070909	Oversensing		
									A071204	No Pacing		
							F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A070909	Oversensing		
									A071204	No Pacing		
							F11	Non-Serious Injury/ Illness/ Impairment	A070909	Oversensing		
									A071204	No Pacing		
					CT983	Surgical invasive	E0112	Dizziness	A070909	Oversensing		
									A071204	No Pacing		
							F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A070909	Oversensing		
									A071204	No Pacing		
							F11	Non-Serious Injury/ Illness/ Impairment	A070909	Oversensing		
									A071204	No Pacing		
11184	2021	All other reportable incidents	47265	Dual-chamber implantable pacemaker, rate-responsive	CT141	Electrodes/Leads and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT145	Cardiovascular devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT233	Surgical	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT2406	Implantable	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT2407	Active implantable	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT2727	Active implant devices/ancillaries	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT2750	Implantable cardiac electrical stimulators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT2985	Active	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT3001	Implantable metal	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT3002	Implantable synthetic polymer	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT3006	Implantable inorganic material	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT314	Electrically powered	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT321	Long-term surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT449	Generators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT981	Single use	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT983	Surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
11185	2021	All other reportable incidents	47265	Dual-chamber implantable pacemaker, rate-responsive	CT141	Electrodes/Leads and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	03/06/2021	03/06/2021
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	03/06/2021	03/06/2021

							F19	Surgical Intervention	A070504	Premature Discharge of Battery	03/06/2021	03/06/2021
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	03/06/2021	03/06/2021
					CT145	Cardiovascular devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	03/06/2021	03/06/2021
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	03/06/2021	03/06/2021
							F19	Surgical Intervention	A070504	Premature Discharge of Battery	03/06/2021	03/06/2021
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	03/06/2021	03/06/2021
					CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	03/06/2021	03/06/2021
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	03/06/2021	03/06/2021
							F19	Surgical Intervention	A070504	Premature Discharge of Battery	03/06/2021	03/06/2021
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	03/06/2021	03/06/2021
					CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	03/06/2021	03/06/2021
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	03/06/2021	03/06/2021
							F19	Surgical Intervention	A070504	Premature Discharge of Battery	03/06/2021	03/06/2021
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	03/06/2021	03/06/2021
					CT233	Surgical	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	03/06/2021	03/06/2021
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	03/06/2021	03/06/2021
							F19	Surgical Intervention	A070504	Premature Discharge of Battery	03/06/2021	03/06/2021
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	03/06/2021	03/06/2021
					CT2406	Implantable	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	03/06/2021	03/06/2021
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	03/06/2021	03/06/2021
							F19	Surgical Intervention	A070504	Premature Discharge of Battery	03/06/2021	03/06/2021
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	03/06/2021	03/06/2021
					CT2407	Active implantable	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	03/06/2021	03/06/2021
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	03/06/2021	03/06/2021
							F19	Surgical Intervention	A070504	Premature Discharge of Battery	03/06/2021	03/06/2021
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	03/06/2021	03/06/2021
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	03/06/2021	03/06/2021
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	03/06/2021	03/06/2021
							F19	Surgical Intervention	A070504	Premature Discharge of Battery	03/06/2021	03/06/2021
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	03/06/2021	03/06/2021
					CT2727	Active implant devices/ancillaries	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	03/06/2021	03/06/2021
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	03/06/2021	03/06/2021
							F19	Surgical Intervention	A070504	Premature Discharge of Battery	03/06/2021	03/06/2021
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	03/06/2021	03/06/2021
					CT2750	Implantable cardiac electrical stimulators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	03/06/2021	03/06/2021
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	03/06/2021	03/06/2021
							F19	Surgical Intervention	A070504	Premature Discharge of Battery	03/06/2021	03/06/2021
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	03/06/2021	03/06/2021
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	03/06/2021	03/06/2021
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	03/06/2021	03/06/2021
							F19	Surgical Intervention	A070504	Premature Discharge of Battery	03/06/2021	03/06/2021
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	03/06/2021	03/06/2021
					CT2985	Active	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	03/06/2021	03/06/2021
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	03/06/2021	03/06/2021
							F19	Surgical Intervention	A070504	Premature Discharge of Battery	03/06/2021	03/06/2021
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	03/06/2021	03/06/2021
					CT3001	Implantable metal	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	03/06/2021	03/06/2021
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	03/06/2021	03/06/2021
							F19	Surgical Intervention	A070504	Premature Discharge of Battery	03/06/2021	03/06/2021
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	03/06/2021	03/06/2021
					CT3002	Implantable synthetic polymer	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	03/06/2021	03/06/2021
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	03/06/2021	03/06/2021
							F19	Surgical Intervention	A070504	Premature Discharge of Battery	03/06/2021	03/06/2021
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	03/06/2021	03/06/2021
					CT3006	Implantable inorganic material	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	03/06/2021	03/06/2021
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	03/06/2021	03/06/2021
							F19	Surgical Intervention	A070504	Premature Discharge of Battery	03/06/2021	03/06/2021
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	03/06/2021	03/06/2021
					CT314	Electrically powered	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	03/06/2021	03/06/2021
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	03/06/2021	03/06/2021
							F19	Surgical Intervention	A070504	Premature Discharge of Battery	03/06/2021	03/06/2021
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	03/06/2021	03/06/2021
					CT321	Long-term surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	03/06/2021	03/06/2021
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	03/06/2021	03/06/2021
							F19	Surgical Intervention	A070504	Premature Discharge of Battery	03/06/2021	03/06/2021
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	03/06/2021	03/06/2021
					CT449	Generators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	03/06/2021	03/06/2021
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	03/06/2021	03/06/2021
							F19	Surgical Intervention	A070504	Premature Discharge of Battery	03/06/2021	03/06/2021
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	03/06/2021	03/06/2021
					CT981	Single use	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	03/06/2021	03/06/2021
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	03/06/2021	03/06/2021
							F19	Surgical Intervention	A070504	Premature Discharge of Battery	03/06/2021	03/06/2021
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	03/06/2021	03/06/2021
					CT983	Surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	03/06/2021	03/06/2021
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	03/06/2021	03/06/2021
							F19	Surgical Intervention	A070504	Premature Discharge of Battery	03/06/2021	03/06/2021
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	03/06/2021	03/06/2021
11186	2021	All other reportable incidents	36241	Pacemaker/defibrillator lead	CT1684	Cardiac pacemakers and associated devices	E0112	Dizziness	A070101	Failure to Capture	18/05/2021	18/05/2021
									A070102	High Capture Threshold	18/05/2021	18/05/2021
							F08	Hospitalization or Prolonged Hospitalization	A070101	Failure to Capture	18/05/2021	18/05/2021
									A070102	High Capture Threshold	18/05/2021	18/05/2021
							F19	Surgical Intervention	A070101	Failure to Capture	18/05/2021	18/05/2021
									A070102	High Capture Threshold	18/05/2021	18/05/2021
							F1905	Device Revision or Replacement	A070101	Failure to Capture	18/05/2021	18/05/2021
									A070102	High Capture Threshold	18/05/2021	18/05/2021
11187	2021	All other reportable incidents	36241	Pacemaker/defibrillator lead	CT1684	Cardiac pacemakers and associated devices	E2104	Electric Shock	A051201	Device Dislodged or Dislocated		
									A071302	Inappropriate/Inadequate Shock/Stimulation		

							E2403	No Clinical Signs, Symptoms or Conditions	A051201	Device Dislodged or Dislocated		
									A071302	Inappropriate/Inadequate Shock/Stimulation		
							F1904	Device Repositioning	A051201	Device Dislodged or Dislocated		
									A071302	Inappropriate/Inadequate Shock/Stimulation		
11188	2021	All other reportable incidents	35852	Single-chamber implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem	22/07/2021	22/07/2021
									A070504	Premature Discharge of Battery	22/07/2021	22/07/2021
							F08	Hospitalization or Prolonged Hospitalization	A0705	Battery Problem	22/07/2021	22/07/2021
									A070504	Premature Discharge of Battery	22/07/2021	22/07/2021
							F1905	Device Revision or Replacement	A0705	Battery Problem	22/07/2021	22/07/2021
									A070504	Premature Discharge of Battery	22/07/2021	22/07/2021
11189	2021	All other reportable incidents	35852	Single-chamber implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E2401	Insufficient Information	A0705	Battery Problem	11/07/2021	11/07/2021
									A070504	Premature Discharge of Battery	11/07/2021	11/07/2021
							E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem	11/07/2021	11/07/2021
									A070504	Premature Discharge of Battery	11/07/2021	11/07/2021
							F24	Insufficient Health Impact Information	A0705	Battery Problem	11/07/2021	11/07/2021
									A070504	Premature Discharge of Battery	11/07/2021	11/07/2021
11190	2021	All other reportable incidents	35852	Single-chamber implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	17/05/2021	17/05/2021
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	17/05/2021	17/05/2021
							F12	Serious Injury/ Illness/ Impairment	A070504	Premature Discharge of Battery	17/05/2021	17/05/2021
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	17/05/2021	17/05/2021
11191	2021	All other reportable incidents	47267	Single-chamber implantable pacemaker, rate-responsive	CT141	Electrodes/Leads and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	25/08/2021	25/08/2021
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	25/08/2021	25/08/2021
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	25/08/2021	25/08/2021
					CT145	Cardiovascular devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	25/08/2021	25/08/2021
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	25/08/2021	25/08/2021
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	25/08/2021	25/08/2021
					CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	25/08/2021	25/08/2021
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	25/08/2021	25/08/2021
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	25/08/2021	25/08/2021
					CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	25/08/2021	25/08/2021
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	25/08/2021	25/08/2021
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	25/08/2021	25/08/2021
					CT233	Surgical	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	25/08/2021	25/08/2021
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	25/08/2021	25/08/2021
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	25/08/2021	25/08/2021
					CT2406	Implantable	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	25/08/2021	25/08/2021
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	25/08/2021	25/08/2021
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	25/08/2021	25/08/2021
					CT2407	Active implantable	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	25/08/2021	25/08/2021
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	25/08/2021	25/08/2021
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	25/08/2021	25/08/2021
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	25/08/2021	25/08/2021
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	25/08/2021	25/08/2021
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	25/08/2021	25/08/2021
					CT2727	Active implant devices/ancillaries	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	25/08/2021	25/08/2021
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	25/08/2021	25/08/2021
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	25/08/2021	25/08/2021
					CT2750	Implantable cardiac electrical stimulators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	25/08/2021	25/08/2021
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	25/08/2021	25/08/2021
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	25/08/2021	25/08/2021
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	25/08/2021	25/08/2021
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	25/08/2021	25/08/2021
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	25/08/2021	25/08/2021
					CT2985	Active	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	25/08/2021	25/08/2021
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	25/08/2021	25/08/2021
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	25/08/2021	25/08/2021
					CT3001	Implantable metal	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	25/08/2021	25/08/2021
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	25/08/2021	25/08/2021
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	25/08/2021	25/08/2021
					CT3002	Implantable synthetic polymer	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	25/08/2021	25/08/2021
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	25/08/2021	25/08/2021
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	25/08/2021	25/08/2021
					CT3006	Implantable inorganic material	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	25/08/2021	25/08/2021
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	25/08/2021	25/08/2021
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	25/08/2021	25/08/2021
					CT314	Electrically powered	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	25/08/2021	25/08/2021
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	25/08/2021	25/08/2021
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	25/08/2021	25/08/2021
					CT321	Long-term surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	25/08/2021	25/08/2021
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	25/08/2021	25/08/2021
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	25/08/2021	25/08/2021
					CT449	Generators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	25/08/2021	25/08/2021
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	25/08/2021	25/08/2021
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	25/08/2021	25/08/2021
					CT981	Single use	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	25/08/2021	25/08/2021
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	25/08/2021	25/08/2021
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	25/08/2021	25/08/2021
					CT983	Surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	25/08/2021	25/08/2021
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	25/08/2021	25/08/2021
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	25/08/2021	25/08/2021
11192	2021	All other reportable incidents	17882	Physiologic-monitoring defibrillation system	CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070908	Failure to Sense		
							F26	No Health Consequences or Impact	A070908	Failure to Sense		
11193	2021	All other reportable incidents	35039	Epicardial pacing lead	CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A072202	Low Impedance		
							F26	No Health Consequences or Impact	A072202	Low Impedance		
11194	2021	All other reportable incidents	35223	Endocardial/interventricular septal pacing lead	CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070909	Oversensing		
									A072202	Low Impedance		
							F26	No Health Consequences or Impact	A070909	Oversensing		
									A072202	Low Impedance		

11195	2021	All other reportable incidents	35223	Endocardial/interventricular septal pacing lead	CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A0722	Impedance Problem		
									A072201	High Impedance		
							F26	No Health Consequences or Impact	A0722	Impedance Problem		
									A072201	High Impedance		
11196	2021	All other reportable incidents	35223	Endocardial/interventricular septal pacing lead	CT1684	Cardiac pacemakers and associated devices	E0604	Cardiac Perforation	A070101	Failure to Capture	07/06/2021	07/06/2021
									A1502	Positioning Problem	07/06/2021	07/06/2021
							E0619	Pericardial Effusion	A070101	Failure to Capture	07/06/2021	07/06/2021
									A1502	Positioning Problem	07/06/2021	07/06/2021
							E2114	Perforation	A070101	Failure to Capture	07/06/2021	07/06/2021
									A1502	Positioning Problem	07/06/2021	07/06/2021
							F12	Serious Injury/ Illness/ Impairment	A070101	Failure to Capture	07/06/2021	07/06/2021
									A1502	Positioning Problem	07/06/2021	07/06/2021
							F1203	Life Threatening Illness or Injury	A070101	Failure to Capture	07/06/2021	07/06/2021
									A1502	Positioning Problem	07/06/2021	07/06/2021
							F15	Recognized Device or Procedural Complication	A070101	Failure to Capture	07/06/2021	07/06/2021
									A1502	Positioning Problem	07/06/2021	07/06/2021
11197	2021	All other reportable incidents	47270	Cardiac resynchronization therapy implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	09/06/2021	09/06/2021
							F1901	Additional Surgery	A070504	Premature Discharge of Battery	09/06/2021	09/06/2021
11198	2021	All other reportable incidents	35852	Single-chamber implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
11199	2021	All other reportable incidents	47265	Dual-chamber implantable pacemaker, rate-responsive	CT141	Electrodes/Leads and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	12/07/2021	12/07/2021
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery	12/07/2021	12/07/2021
					CT145	Cardiovascular devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	12/07/2021	12/07/2021
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery	12/07/2021	12/07/2021
					CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	12/07/2021	12/07/2021
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery	12/07/2021	12/07/2021
					CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	12/07/2021	12/07/2021
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery	12/07/2021	12/07/2021
					CT233	Surgical	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	12/07/2021	12/07/2021
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery	12/07/2021	12/07/2021
					CT2406	Implantable	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	12/07/2021	12/07/2021
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery	12/07/2021	12/07/2021
					CT2407	Active implantable	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	12/07/2021	12/07/2021
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery	12/07/2021	12/07/2021
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	12/07/2021	12/07/2021
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery	12/07/2021	12/07/2021
					CT2727	Active implant devices/ancillaries	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	12/07/2021	12/07/2021
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery	12/07/2021	12/07/2021
					CT2750	Implantable cardiac electrical stimulators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	12/07/2021	12/07/2021
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery	12/07/2021	12/07/2021
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	12/07/2021	12/07/2021
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery	12/07/2021	12/07/2021
					CT2985	Active	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	12/07/2021	12/07/2021
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery	12/07/2021	12/07/2021
					CT3001	Implantable metal	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	12/07/2021	12/07/2021
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery	12/07/2021	12/07/2021
					CT3002	Implantable synthetic polymer	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	12/07/2021	12/07/2021
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery	12/07/2021	12/07/2021
					CT3006	Implantable inorganic material	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	12/07/2021	12/07/2021
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery	12/07/2021	12/07/2021
					CT314	Electrically powered	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	12/07/2021	12/07/2021
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery	12/07/2021	12/07/2021
					CT321	Long-term surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	12/07/2021	12/07/2021
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery	12/07/2021	12/07/2021
					CT449	Generators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	12/07/2021	12/07/2021
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery	12/07/2021	12/07/2021
					CT981	Single use	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	12/07/2021	12/07/2021
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery	12/07/2021	12/07/2021
					CT983	Surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	12/07/2021	12/07/2021
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery	12/07/2021	12/07/2021
11200	2021	All other reportable incidents	47263	Cardiac resynchronization therapy implantable pacemaker	CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A0712	Pacing Problem	01/06/2021	01/06/2021
									A0908	Incorrect, Inadequate or Imprecise Result or Readings	01/06/2021	01/06/2021
							F08	Hospitalization or Prolonged Hospitalization	A0712	Pacing Problem	01/06/2021	01/06/2021
									A0908	Incorrect, Inadequate or Imprecise Result or Readings	01/06/2021	01/06/2021
							F1903	Device Explantation	A0712	Pacing Problem	01/06/2021	01/06/2021
									A0908	Incorrect, Inadequate or Imprecise Result or Readings	01/06/2021	01/06/2021
							F1905	Device Revision or Replacement	A0712	Pacing Problem	01/06/2021	01/06/2021
									A0908	Incorrect, Inadequate or Imprecise Result or Readings	01/06/2021	01/06/2021
					CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A0712	Pacing Problem	01/06/2021	01/06/2021
									A0908	Incorrect, Inadequate or Imprecise Result or Readings	01/06/2021	01/06/2021
							F08	Hospitalization or Prolonged Hospitalization	A0712	Pacing Problem	01/06/2021	01/06/2021
									A0908	Incorrect, Inadequate or Imprecise Result or Readings	01/06/2021	01/06/2021
							F1903	Device Explantation	A0712	Pacing Problem	01/06/2021	01/06/2021
									A0908	Incorrect, Inadequate or Imprecise Result or Readings	01/06/2021	01/06/2021
							F1905	Device Revision or Replacement	A0712	Pacing Problem	01/06/2021	01/06/2021
									A0908	Incorrect, Inadequate or Imprecise Result or Readings	01/06/2021	01/06/2021
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A0712	Pacing Problem	01/06/2021	01/06/2021
									A0908	Incorrect, Inadequate or Imprecise Result or Readings	01/06/2021	01/06/2021
							F08	Hospitalization or Prolonged Hospitalization	A0712	Pacing Problem	01/06/2021	01/06/2021

									A0908	Incorrect, Inadequate or Imprecise Result or Readings	01/06/2021	01/06/2021
							F1903	Device Explantation	A0712	Pacing Problem	01/06/2021	01/06/2021
									A0908	Incorrect, Inadequate or Imprecise Result or Readings	01/06/2021	01/06/2021
							F1905	Device Revision or Replacement	A0712	Pacing Problem	01/06/2021	01/06/2021
									A0908	Incorrect, Inadequate or Imprecise Result or Readings	01/06/2021	01/06/2021
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A0712	Pacing Problem	01/06/2021	01/06/2021
									A0908	Incorrect, Inadequate or Imprecise Result or Readings	01/06/2021	01/06/2021
							F08	Hospitalization or Prolonged Hospitalization	A0712	Pacing Problem	01/06/2021	01/06/2021
									A0908	Incorrect, Inadequate or Imprecise Result or Readings	01/06/2021	01/06/2021
							F1903	Device Explantation	A0712	Pacing Problem	01/06/2021	01/06/2021
									A0908	Incorrect, Inadequate or Imprecise Result or Readings	01/06/2021	01/06/2021
							F1905	Device Revision or Replacement	A0712	Pacing Problem	01/06/2021	01/06/2021
									A0908	Incorrect, Inadequate or Imprecise Result or Readings	01/06/2021	01/06/2021
11201	2021	All other reportable incidents	36241	Pacemaker/defibrillator lead	CT1684	Cardiac pacemakers and associated devices	E2104	Electric Shock	A071302	Inappropriate/Inadequate Shock/Stimulation	10/06/2021	10/06/2021
							F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A071302	Inappropriate/Inadequate Shock/Stimulation	10/06/2021	10/06/2021
							F1905	Device Revision or Replacement	A071302	Inappropriate/Inadequate Shock/Stimulation	10/06/2021	10/06/2021
11202	2021	All other reportable incidents	35852	Single-chamber implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A070909	Oversensing		
							F26	No Health Consequences or Impact	A070909	Oversensing		
11203	2021	All other reportable incidents	35852	Single-chamber implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A070909	Oversensing		
							F26	No Health Consequences or Impact	A070909	Oversensing		
11204	2021	All other reportable incidents	35852	Single-chamber implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A070909	Oversensing		
							F26	No Health Consequences or Impact	A070909	Oversensing		
11205	2021	All other reportable incidents	35852	Single-chamber implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem		
							F26	No Health Consequences or Impact	A0705	Battery Problem		
11206	2021	All other reportable incidents	35852	Single-chamber implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	22/08/2021	22/08/2021
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	22/08/2021	22/08/2021
							F1903	Device Explantation	A070504	Premature Discharge of Battery	22/08/2021	22/08/2021
11207	2021	All other reportable incidents	17882	Physiologic-monitoring defibrillation system	CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070803	Failure to Power Up		
							F27	Problem Identified During Non-Clinical Procedure	A070803	Failure to Power Up		
11208	2021	All other reportable incidents	17882	Physiologic-monitoring defibrillation system	CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A07	Electrical /Electronic Property Problem		
							F26	No Health Consequences or Impact	A07	Electrical /Electronic Property Problem		
11209	2021	All other reportable incidents	35852	Single-chamber implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E2311	Discomfort	A24	No Apparent Device Problem	14/04/2021	14/04/2021
							F1905	Device Revision or Replacement	A24	No Apparent Device Problem	14/04/2021	14/04/2021
11210	2021	Unanticipated serious deterioration in state of health	35223	Endocardial/interventricular septal pacing lead	CT1684	Cardiac pacemakers and associated devices	E011903	Syncope/Fainting	A070101	Failure to Capture	09/06/2021	09/06/2021
									A070102	High Capture Threshold	09/06/2021	09/06/2021
							E060104	Bradycardia	A070101	Failure to Capture	09/06/2021	09/06/2021
									A070102	High Capture Threshold	09/06/2021	09/06/2021
							E060106	Heart Block	A070101	Failure to Capture	09/06/2021	09/06/2021
									A070102	High Capture Threshold	09/06/2021	09/06/2021
							E0623	Unspecified Heart Problem	A070101	Failure to Capture	09/06/2021	09/06/2021
									A070102	High Capture Threshold	09/06/2021	09/06/2021
							E2104	Electric Shock	A070101	Failure to Capture	09/06/2021	09/06/2021
									A070102	High Capture Threshold	09/06/2021	09/06/2021
							E233601	Cardiogenic Shock	A070101	Failure to Capture	09/06/2021	09/06/2021
									A070102	High Capture Threshold	09/06/2021	09/06/2021
							F1203	Life Threatening Illness or Injury	A070101	Failure to Capture	09/06/2021	09/06/2021
									A070102	High Capture Threshold	09/06/2021	09/06/2021
							F1905	Device Revision or Replacement	A070101	Failure to Capture	09/06/2021	09/06/2021
									A070102	High Capture Threshold	09/06/2021	09/06/2021
							F21	Unexpected Deterioration	A070101	Failure to Capture	09/06/2021	09/06/2021
									A070102	High Capture Threshold	09/06/2021	09/06/2021
11211	2021	All other reportable incidents	47267	Single-chamber implantable pacemaker, rate-responsive	CT141	Electrodes/Leads and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	05/08/2021	05/08/2021
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	05/08/2021	05/08/2021
							F1903	Device Explantation	A070504	Premature Discharge of Battery	05/08/2021	05/08/2021
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	05/08/2021	05/08/2021
					CT145	Cardiovascular devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	05/08/2021	05/08/2021
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	05/08/2021	05/08/2021
							F1903	Device Explantation	A070504	Premature Discharge of Battery	05/08/2021	05/08/2021
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	05/08/2021	05/08/2021
					CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	05/08/2021	05/08/2021
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	05/08/2021	05/08/2021
							F1903	Device Explantation	A070504	Premature Discharge of Battery	05/08/2021	05/08/2021
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	05/08/2021	05/08/2021
					CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	05/08/2021	05/08/2021
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	05/08/2021	05/08/2021
							F1903	Device Explantation	A070504	Premature Discharge of Battery	05/08/2021	05/08/2021
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	05/08/2021	05/08/2021
					CT233	Surgical	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	05/08/2021	05/08/2021
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	05/08/2021	05/08/2021
							F1903	Device Explantation	A070504	Premature Discharge of Battery	05/08/2021	05/08/2021
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	05/08/2021	05/08/2021
					CT2406	Implantable	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	05/08/2021	05/08/2021
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	05/08/2021	05/08/2021
							F1903	Device Explantation	A070504	Premature Discharge of Battery	05/08/2021	05/08/2021
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	05/08/2021	05/08/2021
					CT2407	Active implantable	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	05/08/2021	05/08/2021
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	05/08/2021	05/08/2021
							F1903	Device Explantation	A070504	Premature Discharge of Battery	05/08/2021	05/08/2021
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	05/08/2021	05/08/2021
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	05/08/2021	05/08/2021
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	05/08/2021	05/08/2021
							F1903	Device Explantation	A070504	Premature Discharge of Battery	05/08/2021	05/08/2021

							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	05/08/2021	05/08/2021
					CT2727	Active implant devices/ancillaries	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	05/08/2021	05/08/2021
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	05/08/2021	05/08/2021
							F1903	Device Explantation	A070504	Premature Discharge of Battery	05/08/2021	05/08/2021
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	05/08/2021	05/08/2021
					CT2750	Implantable cardiac electrical stimulators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	05/08/2021	05/08/2021
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	05/08/2021	05/08/2021
							F1903	Device Explantation	A070504	Premature Discharge of Battery	05/08/2021	05/08/2021
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	05/08/2021	05/08/2021
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	05/08/2021	05/08/2021
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	05/08/2021	05/08/2021
							F1903	Device Explantation	A070504	Premature Discharge of Battery	05/08/2021	05/08/2021
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	05/08/2021	05/08/2021
					CT2985	Active	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	05/08/2021	05/08/2021
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	05/08/2021	05/08/2021
							F1903	Device Explantation	A070504	Premature Discharge of Battery	05/08/2021	05/08/2021
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	05/08/2021	05/08/2021
					CT3001	Implantable metal	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	05/08/2021	05/08/2021
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	05/08/2021	05/08/2021
							F1903	Device Explantation	A070504	Premature Discharge of Battery	05/08/2021	05/08/2021
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	05/08/2021	05/08/2021
					CT3002	Implantable synthetic polymer	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	05/08/2021	05/08/2021
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	05/08/2021	05/08/2021
							F1903	Device Explantation	A070504	Premature Discharge of Battery	05/08/2021	05/08/2021
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	05/08/2021	05/08/2021
					CT3006	Implantable inorganic material	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	05/08/2021	05/08/2021
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	05/08/2021	05/08/2021
							F1903	Device Explantation	A070504	Premature Discharge of Battery	05/08/2021	05/08/2021
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	05/08/2021	05/08/2021
					CT314	Electrically powered	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	05/08/2021	05/08/2021
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	05/08/2021	05/08/2021
							F1903	Device Explantation	A070504	Premature Discharge of Battery	05/08/2021	05/08/2021
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	05/08/2021	05/08/2021
					CT321	Long-term surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	05/08/2021	05/08/2021
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	05/08/2021	05/08/2021
							F1903	Device Explantation	A070504	Premature Discharge of Battery	05/08/2021	05/08/2021
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	05/08/2021	05/08/2021
					CT449	Generators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	05/08/2021	05/08/2021
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	05/08/2021	05/08/2021
							F1903	Device Explantation	A070504	Premature Discharge of Battery	05/08/2021	05/08/2021
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	05/08/2021	05/08/2021
					CT981	Single use	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	05/08/2021	05/08/2021
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	05/08/2021	05/08/2021
							F1903	Device Explantation	A070504	Premature Discharge of Battery	05/08/2021	05/08/2021
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	05/08/2021	05/08/2021
					CT983	Surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	05/08/2021	05/08/2021
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	05/08/2021	05/08/2021
							F1903	Device Explantation	A070504	Premature Discharge of Battery	05/08/2021	05/08/2021
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	05/08/2021	05/08/2021
11212	2021	All other reportable incidents	47270	Cardiac resynchronization therapy implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A072201	High Impedance		
							F26	No Health Consequences or Impact	A072201	High Impedance		
11213	2021	All other reportable incidents	36241	Pacemaker/defibrillator lead	CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A072201	High Impedance		
							F26	No Health Consequences or Impact	A072201	High Impedance		
11214	2021	All other reportable incidents	35852	Single-chamber implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
11215	2021	All other reportable incidents	35852	Single-chamber implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E2104	Electric Shock	A070909	Oversensing		
									A071302	Inappropriate/Inadequate Shock/Stimulation		
									A090801	Signal Artifact/Noise		
							E2403	No Clinical Signs, Symptoms or Conditions	A070909	Oversensing		
									A071302	Inappropriate/Inadequate Shock/Stimulation		
									A090801	Signal Artifact/Noise		
							F19	Surgical Intervention	A070909	Oversensing		
									A071302	Inappropriate/Inadequate Shock/Stimulation		
									A090801	Signal Artifact/Noise		
11216	2021	All other reportable incidents	17882	Physiologic-monitoring defibrillation system	CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070803	Failure to Power Up		
							F27	Problem Identified During Non-Clinical Procedure	A070803	Failure to Power Up		
11217	2021	All other reportable incidents	35039	Epicardial pacing lead	CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A040101	Fracture	09/06/2021	09/06/2021
									A070909	Oversensing	09/06/2021	09/06/2021
									A090801	Signal Artifact/Noise	09/06/2021	09/06/2021
							F1905	Device Revision or Replacement	A040101	Fracture	09/06/2021	09/06/2021
									A070909	Oversensing	09/06/2021	09/06/2021
									A090801	Signal Artifact/Noise	09/06/2021	09/06/2021
							F26	No Health Consequences or Impact	A040101	Fracture	09/06/2021	09/06/2021
									A070909	Oversensing	09/06/2021	09/06/2021
									A090801	Signal Artifact/Noise	09/06/2021	09/06/2021
11218	2021	Unanticipated serious deterioration in state of health	47267	Single-chamber implantable pacemaker, rate-responsive	CT141	Electrodes/Leads and associated devices	E0602	Cardiac Arrest	A070101	Failure to Capture		
							F1203	Life Threatening Illness or Injury	A070101	Failure to Capture		
					CT145	Cardiovascular devices	E0602	Cardiac Arrest	A070101	Failure to Capture		
							F1203	Life Threatening Illness or Injury	A070101	Failure to Capture		
					CT1684	Cardiac pacemakers and associated devices	E0602	Cardiac Arrest	A070101	Failure to Capture		
							F1203	Life Threatening Illness or Injury	A070101	Failure to Capture		
					CT1685	Cardiac pacemakers	E0602	Cardiac Arrest	A070101	Failure to Capture		
							F1203	Life Threatening Illness or Injury	A070101	Failure to Capture		
					CT233	Surgical	E0602	Cardiac Arrest	A070101	Failure to Capture		
							F1203	Life Threatening Illness or Injury	A070101	Failure to Capture		
					CT2406	Implantable	E0602	Cardiac Arrest	A070101	Failure to Capture		
							F1203	Life Threatening Illness or Injury	A070101	Failure to Capture		
					CT2407	Active implantable	E0602	Cardiac Arrest	A070101	Failure to Capture		

						F1203	Life Threatening Illness or Injury	A070101	Failure to Capture		
				CT257	Implantable cardiac pacemakers	E0602	Cardiac Arrest	A070101	Failure to Capture		
						F1203	Life Threatening Illness or Injury	A070101	Failure to Capture		
				CT2727	Active implant devices/ancillaries	E0602	Cardiac Arrest	A070101	Failure to Capture		
						F1203	Life Threatening Illness or Injury	A070101	Failure to Capture		
				CT2750	Implantable cardiac electrical stimulators and associated devices	E0602	Cardiac Arrest	A070101	Failure to Capture		
						F1203	Life Threatening Illness or Injury	A070101	Failure to Capture		
				CT2866	Implantable cardiac electrical stimulators	E0602	Cardiac Arrest	A070101	Failure to Capture		
						F1203	Life Threatening Illness or Injury	A070101	Failure to Capture		
				CT2985	Active	E0602	Cardiac Arrest	A070101	Failure to Capture		
						F1203	Life Threatening Illness or Injury	A070101	Failure to Capture		
				CT3001	Implantable metal	E0602	Cardiac Arrest	A070101	Failure to Capture		
						F1203	Life Threatening Illness or Injury	A070101	Failure to Capture		
				CT3002	Implantable synthetic polymer	E0602	Cardiac Arrest	A070101	Failure to Capture		
						F1203	Life Threatening Illness or Injury	A070101	Failure to Capture		
				CT3006	Implantable inorganic material	E0602	Cardiac Arrest	A070101	Failure to Capture		
						F1203	Life Threatening Illness or Injury	A070101	Failure to Capture		
				CT314	Electrically powered	E0602	Cardiac Arrest	A070101	Failure to Capture		
						F1203	Life Threatening Illness or Injury	A070101	Failure to Capture		
				CT321	Long-term surgical invasive	E0602	Cardiac Arrest	A070101	Failure to Capture		
						F1203	Life Threatening Illness or Injury	A070101	Failure to Capture		
				CT449	Generators and associated devices	E0602	Cardiac Arrest	A070101	Failure to Capture		
						F1203	Life Threatening Illness or Injury	A070101	Failure to Capture		
				CT981	Single use	E0602	Cardiac Arrest	A070101	Failure to Capture		
						F1203	Life Threatening Illness or Injury	A070101	Failure to Capture		
				CT983	Surgical invasive	E0602	Cardiac Arrest	A070101	Failure to Capture		
						F1203	Life Threatening Illness or Injury	A070101	Failure to Capture		
11219	2021	All other reportable incidents	47267	Single-chamber implantable pacemaker, rate-responsive	CT141	Electrodes/Leads and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
					CT145	Cardiovascular devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
					CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
					CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
					CT233	Surgical	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
					CT2406	Implantable	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
					CT2407	Active implantable	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
					CT2727	Active implant devices/ancillaries	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
					CT2750	Implantable cardiac electrical stimulators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
					CT2985	Active	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
					CT3001	Implantable metal	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
					CT3002	Implantable synthetic polymer	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
					CT3006	Implantable inorganic material	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
					CT314	Electrically powered	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
					CT321	Long-term surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
					CT449	Generators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
					CT981	Single use	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
					CT983	Surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
11220	2021	All other reportable incidents	47263	Cardiac resynchronization therapy implantable pacemaker	CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A0908	Incorrect, Inadequate or Imprecise Result or Readings	11/06/2021	11/06/2021
							F08	Hospitalization or Prolonged Hospitalization	A0908	Incorrect, Inadequate or Imprecise Result or Readings	11/06/2021	11/06/2021
							F19	Surgical Intervention	A0908	Incorrect, Inadequate or Imprecise Result or Readings	11/06/2021	11/06/2021
							F1905	Device Revision or Replacement	A0908	Incorrect, Inadequate or Imprecise Result or Readings	11/06/2021	11/06/2021
					CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A0908	Incorrect, Inadequate or Imprecise Result or Readings	11/06/2021	11/06/2021
							F08	Hospitalization or Prolonged Hospitalization	A0908	Incorrect, Inadequate or Imprecise Result or Readings	11/06/2021	11/06/2021
							F19	Surgical Intervention	A0908	Incorrect, Inadequate or Imprecise Result or Readings	11/06/2021	11/06/2021
							F1905	Device Revision or Replacement	A0908	Incorrect, Inadequate or Imprecise Result or Readings	11/06/2021	11/06/2021
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A0908	Incorrect, Inadequate or Imprecise Result or Readings	11/06/2021	11/06/2021
							F08	Hospitalization or Prolonged Hospitalization	A0908	Incorrect, Inadequate or Imprecise Result or Readings	11/06/2021	11/06/2021
							F19	Surgical Intervention	A0908	Incorrect, Inadequate or Imprecise Result or Readings	11/06/2021	11/06/2021
							F1905	Device Revision or Replacement	A0908	Incorrect, Inadequate or Imprecise Result or Readings	11/06/2021	11/06/2021
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A0908	Incorrect, Inadequate or Imprecise Result or Readings	11/06/2021	11/06/2021
							F08	Hospitalization or Prolonged Hospitalization	A0908	Incorrect, Inadequate or Imprecise Result or Readings	11/06/2021	11/06/2021
							F19	Surgical Intervention	A0908	Incorrect, Inadequate or Imprecise Result or Readings	11/06/2021	11/06/2021
							F1905	Device Revision or Replacement	A0908	Incorrect, Inadequate or Imprecise Result or Readings	11/06/2021	11/06/2021
11221	2021	All other reportable incidents	47265	Dual-chamber implantable pacemaker, rate-responsive	CT141	Electrodes/Leads and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT145	Cardiovascular devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT233	Surgical	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT2406	Implantable	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		

					CT2407	Active implantable	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT2727	Active implant devices/ancillaries	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT2750	Implantable cardiac electrical stimulators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT2985	Active	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT3001	Implantable metal	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT3002	Implantable synthetic polymer	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT3006	Implantable inorganic material	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT314	Electrically powered	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT321	Long-term surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT449	Generators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT981	Single use	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT983	Surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
11222	2021	All other reportable incidents	35223	Endocardial/interventricular septal pacing lead	CT1684	Cardiac pacemakers and associated devices	E0619	Pericardial Effusion	A27	Appropriate Device Problem Term/Code not Available		
							E0623	Unspecified Heart Problem	A27	Appropriate Device Problem Term/Code not Available		
							F08	Hospitalization or Prolonged Hospitalization	A27	Appropriate Device Problem Term/Code not Available		
							F1901	Additional Surgery	A27	Appropriate Device Problem Term/Code not Available		
							F1905	Device Revision or Replacement	A27	Appropriate Device Problem Term/Code not Available		
							F23	Unexpected Medical Intervention	A27	Appropriate Device Problem Term/Code not Available		
11223	2021	All other reportable incidents	35852	Single-chamber implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A070909	Oversensing		
							F26	No Health Consequences or Impact	A070909	Oversensing		
11224	2021	All other reportable incidents	35852	Single-chamber implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E0601	Arrhythmia	A0711	Interrogation Problem	03/06/2021	03/06/2021
									A071201	Failure to Convert Rhythm	03/06/2021	03/06/2021
									A090801	Signal Artifact/Noise	03/06/2021	03/06/2021
							E0623	Unspecified Heart Problem	A0711	Interrogation Problem	03/06/2021	03/06/2021
									A071201	Failure to Convert Rhythm	03/06/2021	03/06/2021
									A090801	Signal Artifact/Noise	03/06/2021	03/06/2021
							F19	Surgical Intervention	A0711	Interrogation Problem	03/06/2021	03/06/2021
									A071201	Failure to Convert Rhythm	03/06/2021	03/06/2021
									A090801	Signal Artifact/Noise	03/06/2021	03/06/2021
							F1905	Device Revision or Replacement	A0711	Interrogation Problem	03/06/2021	03/06/2021
									A071201	Failure to Convert Rhythm	03/06/2021	03/06/2021
									A090801	Signal Artifact/Noise	03/06/2021	03/06/2021
11225	2021	All other reportable incidents	47267	Single-chamber implantable pacemaker, rate-responsive	CT141	Electrodes/Leads and associated devices	E2007	Fall	A070504	Premature Discharge of Battery		
							F11	Non-Serious Injury/ Illness/ Impairment	A070504	Premature Discharge of Battery		
					CT145	Cardiovascular devices	E2007	Fall	A070504	Premature Discharge of Battery		
							F11	Non-Serious Injury/ Illness/ Impairment	A070504	Premature Discharge of Battery		
					CT1684	Cardiac pacemakers and associated devices	E2007	Fall	A070504	Premature Discharge of Battery		
							F11	Non-Serious Injury/ Illness/ Impairment	A070504	Premature Discharge of Battery		
					CT1685	Cardiac pacemakers	E2007	Fall	A070504	Premature Discharge of Battery		
							F11	Non-Serious Injury/ Illness/ Impairment	A070504	Premature Discharge of Battery		
					CT233	Surgical	E2007	Fall	A070504	Premature Discharge of Battery		
							F11	Non-Serious Injury/ Illness/ Impairment	A070504	Premature Discharge of Battery		
					CT2406	Implantable	E2007	Fall	A070504	Premature Discharge of Battery		
							F11	Non-Serious Injury/ Illness/ Impairment	A070504	Premature Discharge of Battery		
					CT2407	Active implantable	E2007	Fall	A070504	Premature Discharge of Battery		
							F11	Non-Serious Injury/ Illness/ Impairment	A070504	Premature Discharge of Battery		
					CT257	Implantable cardiac pacemakers	E2007	Fall	A070504	Premature Discharge of Battery		
							F11	Non-Serious Injury/ Illness/ Impairment	A070504	Premature Discharge of Battery		
					CT2727	Active implant devices/ancillaries	E2007	Fall	A070504	Premature Discharge of Battery		
							F11	Non-Serious Injury/ Illness/ Impairment	A070504	Premature Discharge of Battery		
					CT2750	Implantable cardiac electrical stimulators and associated devices	E2007	Fall	A070504	Premature Discharge of Battery		
							F11	Non-Serious Injury/ Illness/ Impairment	A070504	Premature Discharge of Battery		
					CT2866	Implantable cardiac electrical stimulators	E2007	Fall	A070504	Premature Discharge of Battery		
							F11	Non-Serious Injury/ Illness/ Impairment	A070504	Premature Discharge of Battery		
					CT2985	Active	E2007	Fall	A070504	Premature Discharge of Battery		
							F11	Non-Serious Injury/ Illness/ Impairment	A070504	Premature Discharge of Battery		
					CT3001	Implantable metal	E2007	Fall	A070504	Premature Discharge of Battery		
							F11	Non-Serious Injury/ Illness/ Impairment	A070504	Premature Discharge of Battery		
					CT3002	Implantable synthetic polymer	E2007	Fall	A070504	Premature Discharge of Battery		
							F11	Non-Serious Injury/ Illness/ Impairment	A070504	Premature Discharge of Battery		
					CT3006	Implantable inorganic material	E2007	Fall	A070504	Premature Discharge of Battery		
							F11	Non-Serious Injury/ Illness/ Impairment	A070504	Premature Discharge of Battery		
					CT314	Electrically powered	E2007	Fall	A070504	Premature Discharge of Battery		
							F11	Non-Serious Injury/ Illness/ Impairment	A070504	Premature Discharge of Battery		
					CT321	Long-term surgical invasive	E2007	Fall	A070504	Premature Discharge of Battery		
							F11	Non-Serious Injury/ Illness/ Impairment	A070504	Premature Discharge of Battery		
					CT449	Generators and associated devices	E2007	Fall	A070504	Premature Discharge of Battery		

							F11	Non-Serious Injury/ Illness/ Impairment	A070504	Premature Discharge of Battery		
					CT981	Single use	E2007	Fall	A070504	Premature Discharge of Battery		
							F11	Non-Serious Injury/ Illness/ Impairment	A070504	Premature Discharge of Battery		
					CT983	Surgical invasive	E2007	Fall	A070504	Premature Discharge of Battery		
							F11	Non-Serious Injury/ Illness/ Impairment	A070504	Premature Discharge of Battery		
11226	2021	All other reportable incidents	35039	Epicardial pacing lead	CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070909	Oversensing	09/06/2021	09/06/2021
									A090801	Signal Artifact/Noise	09/06/2021	09/06/2021
							F1905	Device Revision or Replacement	A070909	Oversensing	09/06/2021	09/06/2021
									A090801	Signal Artifact/Noise	09/06/2021	09/06/2021
							F26	No Health Consequences or Impact	A070909	Oversensing	09/06/2021	09/06/2021
									A090801	Signal Artifact/Noise	09/06/2021	09/06/2021
11227	2021	All other reportable incidents	47270	Cardiac resynchronization therapy implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
11228	2021	All other reportable incidents	35223	Endocardial/interventricular septal pacing lead	CT1684	Cardiac pacemakers and associated devices	E0112	Dizziness	A070101	Failure to Capture	10/06/2021	10/06/2021
							F1901	Additional Surgery	A070101	Failure to Capture	10/06/2021	10/06/2021
							F1905	Device Revision or Replacement	A070101	Failure to Capture	10/06/2021	10/06/2021
							F21	Unexpected Deterioration	A070101	Failure to Capture	10/06/2021	10/06/2021
11229	2021	All other reportable incidents	35852	Single-chamber implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A070909	Oversensing		
									A071302	Inappropriate/Inadequate Shock/Stimulation		
									A090801	Signal Artifact/Noise		
							F26	No Health Consequences or Impact	A070909	Oversensing		
									A071302	Inappropriate/Inadequate Shock/Stimulation		
									A090801	Signal Artifact/Noise		
11230	2021	All other reportable incidents	37265	Dual-chamber implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A072201	High Impedance		
							F26	No Health Consequences or Impact	A072201	High Impedance		
11231	2021	All other reportable incidents	35852	Single-chamber implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E2104	Electric Shock	A0709	Device Sensing Problem		
									A071302	Inappropriate/Inadequate Shock/Stimulation		
							F26	No Health Consequences or Impact	A0709	Device Sensing Problem		
									A071302	Inappropriate/Inadequate Shock/Stimulation		
11232	2021	All other reportable incidents	35223	Endocardial/interventricular septal pacing lead	CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070909	Oversensing	01/06/2021	30/06/2021
									A090801	Signal Artifact/Noise	01/06/2021	30/06/2021
							F26	No Health Consequences or Impact	A070909	Oversensing	01/06/2021	30/06/2021
									A090801	Signal Artifact/Noise	01/06/2021	30/06/2021
11233	2021	All other reportable incidents	35223	Endocardial/interventricular septal pacing lead	CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A0401	Break		
									A070909	Oversensing		
									A072202	Low Impedance		
							F1901	Additional Surgery	A0401	Break		
									A070909	Oversensing		
									A072202	Low Impedance		
11234	2021	All other reportable incidents	35223	Endocardial/interventricular septal pacing lead	CT1684	Cardiac pacemakers and associated devices	E0717	Dyspnea	A040507	Naturally Worn		
									A070101	Failure to Capture		
									A070910	Undersensing		
									A072202	Low Impedance		
							E0752	Unspecified Respiratory Problem	A040507	Naturally Worn		
									A070101	Failure to Capture		
									A070910	Undersensing		
									A072202	Low Impedance		
							F23	Unexpected Medical Intervention	A040507	Naturally Worn		
									A070101	Failure to Capture		
									A070910	Undersensing		
									A072202	Low Impedance		
11235	2021	All other reportable incidents	35223	Endocardial/interventricular septal pacing lead	CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A0401	Break		
									A070101	Failure to Capture		
									A0709	Device Sensing Problem		
									A090801	Signal Artifact/Noise		
							F23	Unexpected Medical Intervention	A0401	Break		
									A070101	Failure to Capture		
									A0709	Device Sensing Problem		
									A090801	Signal Artifact/Noise		
11236	2021	All other reportable incidents	35852	Single-chamber implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E2104	Electric Shock	A071302	Inappropriate/Inadequate Shock/Stimulation		
							E2403	No Clinical Signs, Symptoms or Conditions	A071302	Inappropriate/Inadequate Shock/Stimulation		
							F26	No Health Consequences or Impact	A071302	Inappropriate/Inadequate Shock/Stimulation		
11237	2021	All other reportable incidents	35852	Single-chamber implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem	08/07/2021	08/07/2021
									A070504	Premature Discharge of Battery	08/07/2021	08/07/2021
							F08	Hospitalization or Prolonged Hospitalization	A0705	Battery Problem	08/07/2021	08/07/2021
									A070504	Premature Discharge of Battery	08/07/2021	08/07/2021
							F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A0705	Battery Problem	08/07/2021	08/07/2021
									A070504	Premature Discharge of Battery	08/07/2021	08/07/2021
							F1905	Device Revision or Replacement	A0705	Battery Problem	08/07/2021	08/07/2021
									A070504	Premature Discharge of Battery	08/07/2021	08/07/2021
11238	2021	All other reportable incidents	35852	Single-chamber implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem	07/07/2021	07/07/2021
									A070504	Premature Discharge of Battery	07/07/2021	07/07/2021
							F08	Hospitalization or Prolonged Hospitalization	A0705	Battery Problem	07/07/2021	07/07/2021
									A070504	Premature Discharge of Battery	07/07/2021	07/07/2021
							F1905	Device Revision or Replacement	A0705	Battery Problem	07/07/2021	07/07/2021
									A070504	Premature Discharge of Battery	07/07/2021	07/07/2021
11239	2021	All other reportable incidents	35852	Single-chamber implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem	14/07/2021	14/07/2021
									A070504	Premature Discharge of Battery	14/07/2021	14/07/2021
							F08	Hospitalization or Prolonged Hospitalization	A0705	Battery Problem	14/07/2021	14/07/2021
									A070504	Premature Discharge of Battery	14/07/2021	14/07/2021
							F1905	Device Revision or Replacement	A0705	Battery Problem	14/07/2021	14/07/2021
									A070504	Premature Discharge of Battery	14/07/2021	14/07/2021
11240	2021	All other reportable incidents	36241	Pacemaker/defibrillator lead	CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A072201	High Impedance		
							F26	No Health Consequences or Impact	A072201	High Impedance		
11241	2021	All other reportable incidents	35852	Single-chamber implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A070909	Oversensing	01/03/2020	01/03/2020
							F26	No Health Consequences or Impact	A070909	Oversensing	01/03/2020	01/03/2020
11242	2021	All other reportable incidents	35852	Single-chamber implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A070909	Oversensing		
							F26	No Health Consequences or Impact	A070909	Oversensing		

11243	2021	All other reportable incidents	36241	Pacemaker/defibrillator lead	CT1684	Cardiac pacemakers and associated devices	E2114	Perforation	A051201	Device Dislodged or Dislocated	15/06/2021	15/06/2021
									A070102	High Capture Threshold	15/06/2021	15/06/2021
									A072202	Low Impedance	15/06/2021	15/06/2021
							F19	Surgical Intervention	A051201	Device Dislodged or Dislocated	15/06/2021	15/06/2021
									A070102	High Capture Threshold	15/06/2021	15/06/2021
									A072202	Low Impedance	15/06/2021	15/06/2021
11244	2021	All other reportable incidents	47267	Single-chamber implantable pacemaker, rate-responsive	CT141	Electrodes/Leads and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem	07/07/2021	07/07/2021
									A070504	Premature Discharge of Battery	07/07/2021	07/07/2021
							F08	Hospitalization or Prolonged Hospitalization	A0705	Battery Problem	07/07/2021	07/07/2021
									A070504	Premature Discharge of Battery	07/07/2021	07/07/2021
							F1903	Device Explantation	A0705	Battery Problem	07/07/2021	07/07/2021
									A070504	Premature Discharge of Battery	07/07/2021	07/07/2021
					CT145	Cardiovascular devices	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem	07/07/2021	07/07/2021
									A070504	Premature Discharge of Battery	07/07/2021	07/07/2021
							F08	Hospitalization or Prolonged Hospitalization	A0705	Battery Problem	07/07/2021	07/07/2021
									A070504	Premature Discharge of Battery	07/07/2021	07/07/2021
							F1903	Device Explantation	A0705	Battery Problem	07/07/2021	07/07/2021
									A070504	Premature Discharge of Battery	07/07/2021	07/07/2021
					CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem	07/07/2021	07/07/2021
									A070504	Premature Discharge of Battery	07/07/2021	07/07/2021
							F08	Hospitalization or Prolonged Hospitalization	A0705	Battery Problem	07/07/2021	07/07/2021
									A070504	Premature Discharge of Battery	07/07/2021	07/07/2021
							F1903	Device Explantation	A0705	Battery Problem	07/07/2021	07/07/2021
									A070504	Premature Discharge of Battery	07/07/2021	07/07/2021
					CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem	07/07/2021	07/07/2021
									A070504	Premature Discharge of Battery	07/07/2021	07/07/2021
							F08	Hospitalization or Prolonged Hospitalization	A0705	Battery Problem	07/07/2021	07/07/2021
									A070504	Premature Discharge of Battery	07/07/2021	07/07/2021
							F1903	Device Explantation	A0705	Battery Problem	07/07/2021	07/07/2021
									A070504	Premature Discharge of Battery	07/07/2021	07/07/2021
					CT233	Surgical	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem	07/07/2021	07/07/2021
									A070504	Premature Discharge of Battery	07/07/2021	07/07/2021
							F08	Hospitalization or Prolonged Hospitalization	A0705	Battery Problem	07/07/2021	07/07/2021
									A070504	Premature Discharge of Battery	07/07/2021	07/07/2021
							F1903	Device Explantation	A0705	Battery Problem	07/07/2021	07/07/2021
									A070504	Premature Discharge of Battery	07/07/2021	07/07/2021
					CT2406	Implantable	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem	07/07/2021	07/07/2021
									A070504	Premature Discharge of Battery	07/07/2021	07/07/2021
							F08	Hospitalization or Prolonged Hospitalization	A0705	Battery Problem	07/07/2021	07/07/2021
									A070504	Premature Discharge of Battery	07/07/2021	07/07/2021
							F1903	Device Explantation	A0705	Battery Problem	07/07/2021	07/07/2021
									A070504	Premature Discharge of Battery	07/07/2021	07/07/2021
					CT2407	Active implantable	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem	07/07/2021	07/07/2021
									A070504	Premature Discharge of Battery	07/07/2021	07/07/2021
							F08	Hospitalization or Prolonged Hospitalization	A0705	Battery Problem	07/07/2021	07/07/2021
									A070504	Premature Discharge of Battery	07/07/2021	07/07/2021
							F1903	Device Explantation	A0705	Battery Problem	07/07/2021	07/07/2021
									A070504	Premature Discharge of Battery	07/07/2021	07/07/2021
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem	07/07/2021	07/07/2021
									A070504	Premature Discharge of Battery	07/07/2021	07/07/2021
							F08	Hospitalization or Prolonged Hospitalization	A0705	Battery Problem	07/07/2021	07/07/2021
									A070504	Premature Discharge of Battery	07/07/2021	07/07/2021
							F1903	Device Explantation	A0705	Battery Problem	07/07/2021	07/07/2021
									A070504	Premature Discharge of Battery	07/07/2021	07/07/2021
					CT2727	Active implant devices/ancillaries	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem	07/07/2021	07/07/2021
									A070504	Premature Discharge of Battery	07/07/2021	07/07/2021
							F08	Hospitalization or Prolonged Hospitalization	A0705	Battery Problem	07/07/2021	07/07/2021
									A070504	Premature Discharge of Battery	07/07/2021	07/07/2021
							F1903	Device Explantation	A0705	Battery Problem	07/07/2021	07/07/2021
									A070504	Premature Discharge of Battery	07/07/2021	07/07/2021
					CT2750	Implantable cardiac electrical stimulators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem	07/07/2021	07/07/2021
									A070504	Premature Discharge of Battery	07/07/2021	07/07/2021
							F08	Hospitalization or Prolonged Hospitalization	A0705	Battery Problem	07/07/2021	07/07/2021
									A070504	Premature Discharge of Battery	07/07/2021	07/07/2021
							F1903	Device Explantation	A0705	Battery Problem	07/07/2021	07/07/2021
									A070504	Premature Discharge of Battery	07/07/2021	07/07/2021
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem	07/07/2021	07/07/2021
									A070504	Premature Discharge of Battery	07/07/2021	07/07/2021
							F08	Hospitalization or Prolonged Hospitalization	A0705	Battery Problem	07/07/2021	07/07/2021
									A070504	Premature Discharge of Battery	07/07/2021	07/07/2021
							F1903	Device Explantation	A0705	Battery Problem	07/07/2021	07/07/2021
									A070504	Premature Discharge of Battery	07/07/2021	07/07/2021
					CT2985	Active	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem	07/07/2021	07/07/2021
									A070504	Premature Discharge of Battery	07/07/2021	07/07/2021
							F08	Hospitalization or Prolonged Hospitalization	A0705	Battery Problem	07/07/2021	07/07/2021
									A070504	Premature Discharge of Battery	07/07/2021	07/07/2021
							F1903	Device Explantation	A0705	Battery Problem	07/07/2021	07/07/2021
									A070504	Premature Discharge of Battery	07/07/2021	07/07/2021
					CT3001	Implantable metal	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem	07/07/2021	07/07/2021
									A070504	Premature Discharge of Battery	07/07/2021	07/07/2021
							F08	Hospitalization or Prolonged Hospitalization	A0705	Battery Problem	07/07/2021	07/07/2021
									A070504	Premature Discharge of Battery	07/07/2021	07/07/2021
							F1903	Device Explantation	A0705	Battery Problem	07/07/2021	07/07/2021
									A070504	Premature Discharge of Battery	07/07/2021	07/07/2021
					CT3002	Implantable synthetic polymer	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem	07/07/2021	07/07/2021
									A070504	Premature Discharge of Battery	07/07/2021	07/07/2021
							F08	Hospitalization or Prolonged Hospitalization	A0705	Battery Problem	07/07/2021	07/07/2021
									A070504	Premature Discharge of Battery	07/07/2021	07/07/2021

							F1903	Device Explantation	A0705	Battery Problem	07/07/2021	07/07/2021
									A070504	Premature Discharge of Battery	07/07/2021	07/07/2021
					CT3006	Implantable inorganic material	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem	07/07/2021	07/07/2021
									A070504	Premature Discharge of Battery	07/07/2021	07/07/2021
							F08	Hospitalization or Prolonged Hospitalization	A0705	Battery Problem	07/07/2021	07/07/2021
									A070504	Premature Discharge of Battery	07/07/2021	07/07/2021
							F1903	Device Explantation	A0705	Battery Problem	07/07/2021	07/07/2021
									A070504	Premature Discharge of Battery	07/07/2021	07/07/2021
					CT314	Electrically powered	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem	07/07/2021	07/07/2021
									A070504	Premature Discharge of Battery	07/07/2021	07/07/2021
							F08	Hospitalization or Prolonged Hospitalization	A0705	Battery Problem	07/07/2021	07/07/2021
									A070504	Premature Discharge of Battery	07/07/2021	07/07/2021
							F1903	Device Explantation	A0705	Battery Problem	07/07/2021	07/07/2021
									A070504	Premature Discharge of Battery	07/07/2021	07/07/2021
					CT321	Long-term surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem	07/07/2021	07/07/2021
									A070504	Premature Discharge of Battery	07/07/2021	07/07/2021
							F08	Hospitalization or Prolonged Hospitalization	A0705	Battery Problem	07/07/2021	07/07/2021
									A070504	Premature Discharge of Battery	07/07/2021	07/07/2021
							F1903	Device Explantation	A0705	Battery Problem	07/07/2021	07/07/2021
									A070504	Premature Discharge of Battery	07/07/2021	07/07/2021
					CT449	Generators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem	07/07/2021	07/07/2021
									A070504	Premature Discharge of Battery	07/07/2021	07/07/2021
							F08	Hospitalization or Prolonged Hospitalization	A0705	Battery Problem	07/07/2021	07/07/2021
									A070504	Premature Discharge of Battery	07/07/2021	07/07/2021
							F1903	Device Explantation	A0705	Battery Problem	07/07/2021	07/07/2021
									A070504	Premature Discharge of Battery	07/07/2021	07/07/2021
					CT981	Single use	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem	07/07/2021	07/07/2021
									A070504	Premature Discharge of Battery	07/07/2021	07/07/2021
							F08	Hospitalization or Prolonged Hospitalization	A0705	Battery Problem	07/07/2021	07/07/2021
									A070504	Premature Discharge of Battery	07/07/2021	07/07/2021
							F1903	Device Explantation	A0705	Battery Problem	07/07/2021	07/07/2021
									A070504	Premature Discharge of Battery	07/07/2021	07/07/2021
					CT983	Surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem	07/07/2021	07/07/2021
									A070504	Premature Discharge of Battery	07/07/2021	07/07/2021
							F08	Hospitalization or Prolonged Hospitalization	A0705	Battery Problem	07/07/2021	07/07/2021
									A070504	Premature Discharge of Battery	07/07/2021	07/07/2021
							F1903	Device Explantation	A0705	Battery Problem	07/07/2021	07/07/2021
									A070504	Premature Discharge of Battery	07/07/2021	07/07/2021
11245	2021	All other reportable incidents	37265	Dual-chamber implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem	10/09/2021	10/09/2021
									A070504	Premature Discharge of Battery	10/09/2021	10/09/2021
							F08	Hospitalization or Prolonged Hospitalization	A0705	Battery Problem	10/09/2021	10/09/2021
									A070504	Premature Discharge of Battery	10/09/2021	10/09/2021
							F1905	Device Revision or Replacement	A0705	Battery Problem	10/09/2021	10/09/2021
									A070504	Premature Discharge of Battery	10/09/2021	10/09/2021
11246	2021	All other reportable incidents	47267	Single-chamber implantable pacemaker, rate-responsive	CT141	Electrodes/Leads and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem	29/09/2021	29/09/2021
									A070504	Premature Discharge of Battery	29/09/2021	29/09/2021
							F08	Hospitalization or Prolonged Hospitalization	A0705	Battery Problem	29/09/2021	29/09/2021
									A070504	Premature Discharge of Battery	29/09/2021	29/09/2021
							F1905	Device Revision or Replacement	A0705	Battery Problem	29/09/2021	29/09/2021
									A070504	Premature Discharge of Battery	29/09/2021	29/09/2021
					CT145	Cardiovascular devices	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem	29/09/2021	29/09/2021
									A070504	Premature Discharge of Battery	29/09/2021	29/09/2021
							F08	Hospitalization or Prolonged Hospitalization	A0705	Battery Problem	29/09/2021	29/09/2021
									A070504	Premature Discharge of Battery	29/09/2021	29/09/2021
							F1905	Device Revision or Replacement	A0705	Battery Problem	29/09/2021	29/09/2021
									A070504	Premature Discharge of Battery	29/09/2021	29/09/2021
					CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem	29/09/2021	29/09/2021
									A070504	Premature Discharge of Battery	29/09/2021	29/09/2021
							F08	Hospitalization or Prolonged Hospitalization	A0705	Battery Problem	29/09/2021	29/09/2021
									A070504	Premature Discharge of Battery	29/09/2021	29/09/2021
							F1905	Device Revision or Replacement	A0705	Battery Problem	29/09/2021	29/09/2021
									A070504	Premature Discharge of Battery	29/09/2021	29/09/2021
					CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem	29/09/2021	29/09/2021
									A070504	Premature Discharge of Battery	29/09/2021	29/09/2021
							F08	Hospitalization or Prolonged Hospitalization	A0705	Battery Problem	29/09/2021	29/09/2021
									A070504	Premature Discharge of Battery	29/09/2021	29/09/2021
							F1905	Device Revision or Replacement	A0705	Battery Problem	29/09/2021	29/09/2021
									A070504	Premature Discharge of Battery	29/09/2021	29/09/2021
					CT233	Surgical	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem	29/09/2021	29/09/2021
									A070504	Premature Discharge of Battery	29/09/2021	29/09/2021
							F08	Hospitalization or Prolonged Hospitalization	A0705	Battery Problem	29/09/2021	29/09/2021
									A070504	Premature Discharge of Battery	29/09/2021	29/09/2021
							F1905	Device Revision or Replacement	A0705	Battery Problem	29/09/2021	29/09/2021
									A070504	Premature Discharge of Battery	29/09/2021	29/09/2021
					CT2406	Implantable	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem	29/09/2021	29/09/2021
									A070504	Premature Discharge of Battery	29/09/2021	29/09/2021
							F08	Hospitalization or Prolonged Hospitalization	A0705	Battery Problem	29/09/2021	29/09/2021
									A070504	Premature Discharge of Battery	29/09/2021	29/09/2021
							F1905	Device Revision or Replacement	A0705	Battery Problem	29/09/2021	29/09/2021
									A070504	Premature Discharge of Battery	29/09/2021	29/09/2021
					CT2407	Active implantable	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem	29/09/2021	29/09/2021
									A070504	Premature Discharge of Battery	29/09/2021	29/09/2021
							F08	Hospitalization or Prolonged Hospitalization	A0705	Battery Problem	29/09/2021	29/09/2021
									A070504	Premature Discharge of Battery	29/09/2021	29/09/2021
							F1905	Device Revision or Replacement	A0705	Battery Problem	29/09/2021	29/09/2021
									A070504	Premature Discharge of Battery	29/09/2021	29/09/2021
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem	29/09/2021	29/09/2021
									A070504	Premature Discharge of Battery	29/09/2021	29/09/2021

							F08	Hospitalization or Prolonged Hospitalization	A0705	Battery Problem	29/09/2021	29/09/2021
									A070504	Premature Discharge of Battery	29/09/2021	29/09/2021
							F1905	Device Revision or Replacement	A0705	Battery Problem	29/09/2021	29/09/2021
									A070504	Premature Discharge of Battery	29/09/2021	29/09/2021
					CT2727	Active implant devices/ancillaries	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem	29/09/2021	29/09/2021
									A070504	Premature Discharge of Battery	29/09/2021	29/09/2021
							F08	Hospitalization or Prolonged Hospitalization	A0705	Battery Problem	29/09/2021	29/09/2021
									A070504	Premature Discharge of Battery	29/09/2021	29/09/2021
							F1905	Device Revision or Replacement	A0705	Battery Problem	29/09/2021	29/09/2021
									A070504	Premature Discharge of Battery	29/09/2021	29/09/2021
					CT2750	Implantable cardiac electrical stimulators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem	29/09/2021	29/09/2021
									A070504	Premature Discharge of Battery	29/09/2021	29/09/2021
							F08	Hospitalization or Prolonged Hospitalization	A0705	Battery Problem	29/09/2021	29/09/2021
									A070504	Premature Discharge of Battery	29/09/2021	29/09/2021
							F1905	Device Revision or Replacement	A0705	Battery Problem	29/09/2021	29/09/2021
									A070504	Premature Discharge of Battery	29/09/2021	29/09/2021
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem	29/09/2021	29/09/2021
									A070504	Premature Discharge of Battery	29/09/2021	29/09/2021
							F08	Hospitalization or Prolonged Hospitalization	A0705	Battery Problem	29/09/2021	29/09/2021
									A070504	Premature Discharge of Battery	29/09/2021	29/09/2021
							F1905	Device Revision or Replacement	A0705	Battery Problem	29/09/2021	29/09/2021
									A070504	Premature Discharge of Battery	29/09/2021	29/09/2021
					CT2985	Active	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem	29/09/2021	29/09/2021
									A070504	Premature Discharge of Battery	29/09/2021	29/09/2021
							F08	Hospitalization or Prolonged Hospitalization	A0705	Battery Problem	29/09/2021	29/09/2021
									A070504	Premature Discharge of Battery	29/09/2021	29/09/2021
							F1905	Device Revision or Replacement	A0705	Battery Problem	29/09/2021	29/09/2021
									A070504	Premature Discharge of Battery	29/09/2021	29/09/2021
					CT3001	Implantable metal	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem	29/09/2021	29/09/2021
									A070504	Premature Discharge of Battery	29/09/2021	29/09/2021
							F08	Hospitalization or Prolonged Hospitalization	A0705	Battery Problem	29/09/2021	29/09/2021
									A070504	Premature Discharge of Battery	29/09/2021	29/09/2021
							F1905	Device Revision or Replacement	A0705	Battery Problem	29/09/2021	29/09/2021
									A070504	Premature Discharge of Battery	29/09/2021	29/09/2021
					CT3002	Implantable synthetic polymer	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem	29/09/2021	29/09/2021
									A070504	Premature Discharge of Battery	29/09/2021	29/09/2021
							F08	Hospitalization or Prolonged Hospitalization	A0705	Battery Problem	29/09/2021	29/09/2021
									A070504	Premature Discharge of Battery	29/09/2021	29/09/2021
							F1905	Device Revision or Replacement	A0705	Battery Problem	29/09/2021	29/09/2021
									A070504	Premature Discharge of Battery	29/09/2021	29/09/2021
					CT3006	Implantable inorganic material	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem	29/09/2021	29/09/2021
									A070504	Premature Discharge of Battery	29/09/2021	29/09/2021
							F08	Hospitalization or Prolonged Hospitalization	A0705	Battery Problem	29/09/2021	29/09/2021
									A070504	Premature Discharge of Battery	29/09/2021	29/09/2021
							F1905	Device Revision or Replacement	A0705	Battery Problem	29/09/2021	29/09/2021
									A070504	Premature Discharge of Battery	29/09/2021	29/09/2021
					CT314	Electrically powered	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem	29/09/2021	29/09/2021
									A070504	Premature Discharge of Battery	29/09/2021	29/09/2021
							F08	Hospitalization or Prolonged Hospitalization	A0705	Battery Problem	29/09/2021	29/09/2021
									A070504	Premature Discharge of Battery	29/09/2021	29/09/2021
							F1905	Device Revision or Replacement	A0705	Battery Problem	29/09/2021	29/09/2021
									A070504	Premature Discharge of Battery	29/09/2021	29/09/2021
					CT321	Long-term surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem	29/09/2021	29/09/2021
									A070504	Premature Discharge of Battery	29/09/2021	29/09/2021
							F08	Hospitalization or Prolonged Hospitalization	A0705	Battery Problem	29/09/2021	29/09/2021
									A070504	Premature Discharge of Battery	29/09/2021	29/09/2021
							F1905	Device Revision or Replacement	A0705	Battery Problem	29/09/2021	29/09/2021
									A070504	Premature Discharge of Battery	29/09/2021	29/09/2021
					CT449	Generators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem	29/09/2021	29/09/2021
									A070504	Premature Discharge of Battery	29/09/2021	29/09/2021
							F08	Hospitalization or Prolonged Hospitalization	A0705	Battery Problem	29/09/2021	29/09/2021
									A070504	Premature Discharge of Battery	29/09/2021	29/09/2021
							F1905	Device Revision or Replacement	A0705	Battery Problem	29/09/2021	29/09/2021
									A070504	Premature Discharge of Battery	29/09/2021	29/09/2021
					CT981	Single use	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem	29/09/2021	29/09/2021
									A070504	Premature Discharge of Battery	29/09/2021	29/09/2021
							F08	Hospitalization or Prolonged Hospitalization	A0705	Battery Problem	29/09/2021	29/09/2021
									A070504	Premature Discharge of Battery	29/09/2021	29/09/2021
							F1905	Device Revision or Replacement	A0705	Battery Problem	29/09/2021	29/09/2021
									A070504	Premature Discharge of Battery	29/09/2021	29/09/2021
					CT983	Surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem	29/09/2021	29/09/2021
									A070504	Premature Discharge of Battery	29/09/2021	29/09/2021
							F08	Hospitalization or Prolonged Hospitalization	A0705	Battery Problem	29/09/2021	29/09/2021
									A070504	Premature Discharge of Battery	29/09/2021	29/09/2021
							F1905	Device Revision or Replacement	A0705	Battery Problem	29/09/2021	29/09/2021
									A070504	Premature Discharge of Battery	29/09/2021	29/09/2021
11247	2021	All other reportable incidents	47265	Dual-chamber implantable pacemaker, rate-responsive	CT141	Electrodes/Leads and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem	31/08/2021	31/08/2021
									A070504	Premature Discharge of Battery	31/08/2021	31/08/2021
							F08	Hospitalization or Prolonged Hospitalization	A0705	Battery Problem	31/08/2021	31/08/2021
									A070504	Premature Discharge of Battery	31/08/2021	31/08/2021
							F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A0705	Battery Problem	31/08/2021	31/08/2021
									A070504	Premature Discharge of Battery	31/08/2021	31/08/2021
							F1905	Device Revision or Replacement	A0705	Battery Problem	31/08/2021	31/08/2021
									A070504	Premature Discharge of Battery	31/08/2021	31/08/2021
					CT145	Cardiovascular devices	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem	31/08/2021	31/08/2021
									A070504	Premature Discharge of Battery	31/08/2021	31/08/2021
							F08	Hospitalization or Prolonged Hospitalization	A0705	Battery Problem	31/08/2021	31/08/2021
									A070504	Premature Discharge of Battery	31/08/2021	31/08/2021

							F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A0705	Battery Problem	31/08/2021	31/08/2021
									A070504	Premature Discharge of Battery	31/08/2021	31/08/2021
							F1905	Device Revision or Replacement	A0705	Battery Problem	31/08/2021	31/08/2021
									A070504	Premature Discharge of Battery	31/08/2021	31/08/2021
					CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem	31/08/2021	31/08/2021
									A070504	Premature Discharge of Battery	31/08/2021	31/08/2021
							F08	Hospitalization or Prolonged Hospitalization	A0705	Battery Problem	31/08/2021	31/08/2021
									A070504	Premature Discharge of Battery	31/08/2021	31/08/2021
							F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A0705	Battery Problem	31/08/2021	31/08/2021
									A070504	Premature Discharge of Battery	31/08/2021	31/08/2021
							F1905	Device Revision or Replacement	A0705	Battery Problem	31/08/2021	31/08/2021
									A070504	Premature Discharge of Battery	31/08/2021	31/08/2021
					CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem	31/08/2021	31/08/2021
									A070504	Premature Discharge of Battery	31/08/2021	31/08/2021
							F08	Hospitalization or Prolonged Hospitalization	A0705	Battery Problem	31/08/2021	31/08/2021
									A070504	Premature Discharge of Battery	31/08/2021	31/08/2021
							F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A0705	Battery Problem	31/08/2021	31/08/2021
									A070504	Premature Discharge of Battery	31/08/2021	31/08/2021
							F1905	Device Revision or Replacement	A0705	Battery Problem	31/08/2021	31/08/2021
									A070504	Premature Discharge of Battery	31/08/2021	31/08/2021
					CT233	Surgical	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem	31/08/2021	31/08/2021
									A070504	Premature Discharge of Battery	31/08/2021	31/08/2021
							F08	Hospitalization or Prolonged Hospitalization	A0705	Battery Problem	31/08/2021	31/08/2021
									A070504	Premature Discharge of Battery	31/08/2021	31/08/2021
							F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A0705	Battery Problem	31/08/2021	31/08/2021
									A070504	Premature Discharge of Battery	31/08/2021	31/08/2021
							F1905	Device Revision or Replacement	A0705	Battery Problem	31/08/2021	31/08/2021
									A070504	Premature Discharge of Battery	31/08/2021	31/08/2021
					CT2406	Implantable	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem	31/08/2021	31/08/2021
									A070504	Premature Discharge of Battery	31/08/2021	31/08/2021
							F08	Hospitalization or Prolonged Hospitalization	A0705	Battery Problem	31/08/2021	31/08/2021
									A070504	Premature Discharge of Battery	31/08/2021	31/08/2021
							F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A0705	Battery Problem	31/08/2021	31/08/2021
									A070504	Premature Discharge of Battery	31/08/2021	31/08/2021
							F1905	Device Revision or Replacement	A0705	Battery Problem	31/08/2021	31/08/2021
									A070504	Premature Discharge of Battery	31/08/2021	31/08/2021
					CT2407	Active implantable	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem	31/08/2021	31/08/2021
									A070504	Premature Discharge of Battery	31/08/2021	31/08/2021
							F08	Hospitalization or Prolonged Hospitalization	A0705	Battery Problem	31/08/2021	31/08/2021
									A070504	Premature Discharge of Battery	31/08/2021	31/08/2021
							F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A0705	Battery Problem	31/08/2021	31/08/2021
									A070504	Premature Discharge of Battery	31/08/2021	31/08/2021
							F1905	Device Revision or Replacement	A0705	Battery Problem	31/08/2021	31/08/2021
									A070504	Premature Discharge of Battery	31/08/2021	31/08/2021
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem	31/08/2021	31/08/2021
									A070504	Premature Discharge of Battery	31/08/2021	31/08/2021
							F08	Hospitalization or Prolonged Hospitalization	A0705	Battery Problem	31/08/2021	31/08/2021
									A070504	Premature Discharge of Battery	31/08/2021	31/08/2021
							F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A0705	Battery Problem	31/08/2021	31/08/2021
									A070504	Premature Discharge of Battery	31/08/2021	31/08/2021
							F1905	Device Revision or Replacement	A0705	Battery Problem	31/08/2021	31/08/2021
									A070504	Premature Discharge of Battery	31/08/2021	31/08/2021
					CT2727	Active implant devices/ancillaries	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem	31/08/2021	31/08/2021
									A070504	Premature Discharge of Battery	31/08/2021	31/08/2021
							F08	Hospitalization or Prolonged Hospitalization	A0705	Battery Problem	31/08/2021	31/08/2021
									A070504	Premature Discharge of Battery	31/08/2021	31/08/2021
							F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A0705	Battery Problem	31/08/2021	31/08/2021
									A070504	Premature Discharge of Battery	31/08/2021	31/08/2021
							F1905	Device Revision or Replacement	A0705	Battery Problem	31/08/2021	31/08/2021
									A070504	Premature Discharge of Battery	31/08/2021	31/08/2021
					CT2750	Implantable cardiac electrical stimulators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem	31/08/2021	31/08/2021
									A070504	Premature Discharge of Battery	31/08/2021	31/08/2021
							F08	Hospitalization or Prolonged Hospitalization	A0705	Battery Problem	31/08/2021	31/08/2021
									A070504	Premature Discharge of Battery	31/08/2021	31/08/2021
							F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A0705	Battery Problem	31/08/2021	31/08/2021
									A070504	Premature Discharge of Battery	31/08/2021	31/08/2021
							F1905	Device Revision or Replacement	A0705	Battery Problem	31/08/2021	31/08/2021
									A070504	Premature Discharge of Battery	31/08/2021	31/08/2021
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem	31/08/2021	31/08/2021
									A070504	Premature Discharge of Battery	31/08/2021	31/08/2021
							F08	Hospitalization or Prolonged Hospitalization	A0705	Battery Problem	31/08/2021	31/08/2021
									A070504	Premature Discharge of Battery	31/08/2021	31/08/2021
							F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A0705	Battery Problem	31/08/2021	31/08/2021
									A070504	Premature Discharge of Battery	31/08/2021	31/08/2021
							F1905	Device Revision or Replacement	A0705	Battery Problem	31/08/2021	31/08/2021
									A070504	Premature Discharge of Battery	31/08/2021	31/08/2021
					CT2985	Active	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem	31/08/2021	31/08/2021
									A070504	Premature Discharge of Battery	31/08/2021	31/08/2021
							F08	Hospitalization or Prolonged Hospitalization	A0705	Battery Problem	31/08/2021	31/08/2021
									A070504	Premature Discharge of Battery	31/08/2021	31/08/2021
							F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A0705	Battery Problem	31/08/2021	31/08/2021
									A070504	Premature Discharge of Battery	31/08/2021	31/08/2021
							F1905	Device Revision or Replacement	A0705	Battery Problem	31/08/2021	31/08/2021
									A070504	Premature Discharge of Battery	31/08/2021	31/08/2021
					CT3001	Implantable metal	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem	31/08/2021	31/08/2021
									A070504	Premature Discharge of Battery	31/08/2021	31/08/2021
							F08	Hospitalization or Prolonged Hospitalization	A0705	Battery Problem	31/08/2021	31/08/2021
									A070504	Premature Discharge of Battery	31/08/2021	31/08/2021

							F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A0705	Battery Problem	31/08/2021	31/08/2021
									A070504	Premature Discharge of Battery	31/08/2021	31/08/2021
							F1905	Device Revision or Replacement	A0705	Battery Problem	31/08/2021	31/08/2021
									A070504	Premature Discharge of Battery	31/08/2021	31/08/2021
					CT3002	Implantable synthetic polymer	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem	31/08/2021	31/08/2021
									A070504	Premature Discharge of Battery	31/08/2021	31/08/2021
							F08	Hospitalization or Prolonged Hospitalization	A0705	Battery Problem	31/08/2021	31/08/2021
									A070504	Premature Discharge of Battery	31/08/2021	31/08/2021
							F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A0705	Battery Problem	31/08/2021	31/08/2021
									A070504	Premature Discharge of Battery	31/08/2021	31/08/2021
							F1905	Device Revision or Replacement	A0705	Battery Problem	31/08/2021	31/08/2021
									A070504	Premature Discharge of Battery	31/08/2021	31/08/2021
					CT3006	Implantable inorganic material	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem	31/08/2021	31/08/2021
									A070504	Premature Discharge of Battery	31/08/2021	31/08/2021
							F08	Hospitalization or Prolonged Hospitalization	A0705	Battery Problem	31/08/2021	31/08/2021
									A070504	Premature Discharge of Battery	31/08/2021	31/08/2021
							F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A0705	Battery Problem	31/08/2021	31/08/2021
									A070504	Premature Discharge of Battery	31/08/2021	31/08/2021
							F1905	Device Revision or Replacement	A0705	Battery Problem	31/08/2021	31/08/2021
									A070504	Premature Discharge of Battery	31/08/2021	31/08/2021
					CT314	Electrically powered	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem	31/08/2021	31/08/2021
									A070504	Premature Discharge of Battery	31/08/2021	31/08/2021
							F08	Hospitalization or Prolonged Hospitalization	A0705	Battery Problem	31/08/2021	31/08/2021
									A070504	Premature Discharge of Battery	31/08/2021	31/08/2021
							F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A0705	Battery Problem	31/08/2021	31/08/2021
									A070504	Premature Discharge of Battery	31/08/2021	31/08/2021
							F1905	Device Revision or Replacement	A0705	Battery Problem	31/08/2021	31/08/2021
									A070504	Premature Discharge of Battery	31/08/2021	31/08/2021
					CT321	Long-term surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem	31/08/2021	31/08/2021
									A070504	Premature Discharge of Battery	31/08/2021	31/08/2021
							F08	Hospitalization or Prolonged Hospitalization	A0705	Battery Problem	31/08/2021	31/08/2021
									A070504	Premature Discharge of Battery	31/08/2021	31/08/2021
							F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A0705	Battery Problem	31/08/2021	31/08/2021
									A070504	Premature Discharge of Battery	31/08/2021	31/08/2021
							F1905	Device Revision or Replacement	A0705	Battery Problem	31/08/2021	31/08/2021
									A070504	Premature Discharge of Battery	31/08/2021	31/08/2021
					CT449	Generators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem	31/08/2021	31/08/2021
									A070504	Premature Discharge of Battery	31/08/2021	31/08/2021
							F08	Hospitalization or Prolonged Hospitalization	A0705	Battery Problem	31/08/2021	31/08/2021
									A070504	Premature Discharge of Battery	31/08/2021	31/08/2021
							F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A0705	Battery Problem	31/08/2021	31/08/2021
									A070504	Premature Discharge of Battery	31/08/2021	31/08/2021
							F1905	Device Revision or Replacement	A0705	Battery Problem	31/08/2021	31/08/2021
									A070504	Premature Discharge of Battery	31/08/2021	31/08/2021
					CT981	Single use	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem	31/08/2021	31/08/2021
									A070504	Premature Discharge of Battery	31/08/2021	31/08/2021
							F08	Hospitalization or Prolonged Hospitalization	A0705	Battery Problem	31/08/2021	31/08/2021
									A070504	Premature Discharge of Battery	31/08/2021	31/08/2021
							F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A0705	Battery Problem	31/08/2021	31/08/2021
									A070504	Premature Discharge of Battery	31/08/2021	31/08/2021
							F1905	Device Revision or Replacement	A0705	Battery Problem	31/08/2021	31/08/2021
									A070504	Premature Discharge of Battery	31/08/2021	31/08/2021
					CT983	Surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem	31/08/2021	31/08/2021
									A070504	Premature Discharge of Battery	31/08/2021	31/08/2021
							F08	Hospitalization or Prolonged Hospitalization	A0705	Battery Problem	31/08/2021	31/08/2021
									A070504	Premature Discharge of Battery	31/08/2021	31/08/2021
							F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A0705	Battery Problem	31/08/2021	31/08/2021
									A070504	Premature Discharge of Battery	31/08/2021	31/08/2021
							F1905	Device Revision or Replacement	A0705	Battery Problem	31/08/2021	31/08/2021
									A070504	Premature Discharge of Battery	31/08/2021	31/08/2021
11248	2021	All other reportable incidents	35852	Single-chamber implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E2104	Electric Shock	A071302	Inappropriate/Inadequate Shock/Stimulation		
							F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A071302	Inappropriate/Inadequate Shock/Stimulation		
							F26	No Health Consequences or Impact	A071302	Inappropriate/Inadequate Shock/Stimulation		
11249	2021	All other reportable incidents	47265	Dual-chamber implantable pacemaker, rate-responsive	CT141	Electrodes/Leads and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT145	Cardiovascular devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT233	Surgical	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT2406	Implantable	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT2407	Active implantable	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT2727	Active implant devices/ancillaries	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT2750	Implantable cardiac electrical stimulators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT2985	Active	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT3001	Implantable metal	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		

						F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery			
				CT3002	Implantable synthetic polymer	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery			
						F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery			
				CT3006	Implantable inorganic material	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery			
						F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery			
				CT314	Electrically powered	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery			
						F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery			
				CT321	Long-term surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery			
						F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery			
				CT449	Generators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery			
						F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery			
				CT981	Single use	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery			
						F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery			
				CT983	Surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery			
						F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery			
11250	2021	All other reportable incidents	47267	Single-chamber implantable pacemaker, rate-responsive	CT141	Electrodes/Leads and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT145	Cardiovascular devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT233	Surgical	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT2406	Implantable	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT2407	Active implantable	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT2727	Active implant devices/ancillaries	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT2750	Implantable cardiac electrical stimulators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT2985	Active	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT3001	Implantable metal	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT3002	Implantable synthetic polymer	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT3006	Implantable inorganic material	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT314	Electrically powered	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT321	Long-term surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT449	Generators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT981	Single use	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT983	Surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
11251	2021	All other reportable incidents	47270	Cardiac resynchronization therapy implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A070603	Failure to Charge		
							F23	Unexpected Medical Intervention	A070603	Failure to Charge		
11252	2021	All other reportable incidents	47265	Dual-chamber implantable pacemaker, rate-responsive	CT141	Electrodes/Leads and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem	09/06/2021	09/06/2021
									A070504	Premature Discharge of Battery	09/06/2021	09/06/2021
							F08	Hospitalization or Prolonged Hospitalization	A0705	Battery Problem	09/06/2021	09/06/2021
									A070504	Premature Discharge of Battery	09/06/2021	09/06/2021
							F12	Serious Injury/ Illness/ Impairment	A0705	Battery Problem	09/06/2021	09/06/2021
									A070504	Premature Discharge of Battery	09/06/2021	09/06/2021
							F1903	Device Explantation	A0705	Battery Problem	09/06/2021	09/06/2021
									A070504	Premature Discharge of Battery	09/06/2021	09/06/2021
							F1905	Device Revision or Replacement	A0705	Battery Problem	09/06/2021	09/06/2021
									A070504	Premature Discharge of Battery	09/06/2021	09/06/2021
					CT145	Cardiovascular devices	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem	09/06/2021	09/06/2021
									A070504	Premature Discharge of Battery	09/06/2021	09/06/2021
							F08	Hospitalization or Prolonged Hospitalization	A0705	Battery Problem	09/06/2021	09/06/2021
									A070504	Premature Discharge of Battery	09/06/2021	09/06/2021
							F12	Serious Injury/ Illness/ Impairment	A0705	Battery Problem	09/06/2021	09/06/2021
									A070504	Premature Discharge of Battery	09/06/2021	09/06/2021
							F1903	Device Explantation	A0705	Battery Problem	09/06/2021	09/06/2021
									A070504	Premature Discharge of Battery	09/06/2021	09/06/2021
							F1905	Device Revision or Replacement	A0705	Battery Problem	09/06/2021	09/06/2021
									A070504	Premature Discharge of Battery	09/06/2021	09/06/2021
					CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem	09/06/2021	09/06/2021
									A070504	Premature Discharge of Battery	09/06/2021	09/06/2021
							F08	Hospitalization or Prolonged Hospitalization	A0705	Battery Problem	09/06/2021	09/06/2021
									A070504	Premature Discharge of Battery	09/06/2021	09/06/2021
							F12	Serious Injury/ Illness/ Impairment	A0705	Battery Problem	09/06/2021	09/06/2021
									A070504	Premature Discharge of Battery	09/06/2021	09/06/2021
							F1903	Device Explantation	A0705	Battery Problem	09/06/2021	09/06/2021
									A070504	Premature Discharge of Battery	09/06/2021	09/06/2021
							F1905	Device Revision or Replacement	A0705	Battery Problem	09/06/2021	09/06/2021
									A070504	Premature Discharge of Battery	09/06/2021	09/06/2021
					CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem	09/06/2021	09/06/2021

								A070504	Premature Discharge of Battery	09/06/2021	09/06/2021
						F08	Hospitalization or Prolonged Hospitalization	A0705	Battery Problem	09/06/2021	09/06/2021
								A070504	Premature Discharge of Battery	09/06/2021	09/06/2021
						F12	Serious Injury/ Illness/ Impairment	A0705	Battery Problem	09/06/2021	09/06/2021
								A070504	Premature Discharge of Battery	09/06/2021	09/06/2021
						F1903	Device Explantation	A0705	Battery Problem	09/06/2021	09/06/2021
								A070504	Premature Discharge of Battery	09/06/2021	09/06/2021
						F1905	Device Revision or Replacement	A0705	Battery Problem	09/06/2021	09/06/2021
								A070504	Premature Discharge of Battery	09/06/2021	09/06/2021
					CT233	Surgical	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem	09/06/2021	09/06/2021
									A070504	Premature Discharge of Battery	09/06/2021	09/06/2021
							F08	Hospitalization or Prolonged Hospitalization	A0705	Battery Problem	09/06/2021	09/06/2021
									A070504	Premature Discharge of Battery	09/06/2021	09/06/2021
							F12	Serious Injury/ Illness/ Impairment	A0705	Battery Problem	09/06/2021	09/06/2021
									A070504	Premature Discharge of Battery	09/06/2021	09/06/2021
							F1903	Device Explantation	A0705	Battery Problem	09/06/2021	09/06/2021
									A070504	Premature Discharge of Battery	09/06/2021	09/06/2021
							F1905	Device Revision or Replacement	A0705	Battery Problem	09/06/2021	09/06/2021
									A070504	Premature Discharge of Battery	09/06/2021	09/06/2021
					CT2406	Implantable	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem	09/06/2021	09/06/2021
									A070504	Premature Discharge of Battery	09/06/2021	09/06/2021
							F08	Hospitalization or Prolonged Hospitalization	A0705	Battery Problem	09/06/2021	09/06/2021
									A070504	Premature Discharge of Battery	09/06/2021	09/06/2021
							F12	Serious Injury/ Illness/ Impairment	A0705	Battery Problem	09/06/2021	09/06/2021
									A070504	Premature Discharge of Battery	09/06/2021	09/06/2021
							F1903	Device Explantation	A0705	Battery Problem	09/06/2021	09/06/2021
									A070504	Premature Discharge of Battery	09/06/2021	09/06/2021
							F1905	Device Revision or Replacement	A0705	Battery Problem	09/06/2021	09/06/2021
									A070504	Premature Discharge of Battery	09/06/2021	09/06/2021
					CT2407	Active implantable	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem	09/06/2021	09/06/2021
									A070504	Premature Discharge of Battery	09/06/2021	09/06/2021
							F08	Hospitalization or Prolonged Hospitalization	A0705	Battery Problem	09/06/2021	09/06/2021
									A070504	Premature Discharge of Battery	09/06/2021	09/06/2021
							F12	Serious Injury/ Illness/ Impairment	A0705	Battery Problem	09/06/2021	09/06/2021
									A070504	Premature Discharge of Battery	09/06/2021	09/06/2021
							F1903	Device Explantation	A0705	Battery Problem	09/06/2021	09/06/2021
									A070504	Premature Discharge of Battery	09/06/2021	09/06/2021
							F1905	Device Revision or Replacement	A0705	Battery Problem	09/06/2021	09/06/2021
									A070504	Premature Discharge of Battery	09/06/2021	09/06/2021
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem	09/06/2021	09/06/2021
									A070504	Premature Discharge of Battery	09/06/2021	09/06/2021
							F08	Hospitalization or Prolonged Hospitalization	A0705	Battery Problem	09/06/2021	09/06/2021
									A070504	Premature Discharge of Battery	09/06/2021	09/06/2021
							F12	Serious Injury/ Illness/ Impairment	A0705	Battery Problem	09/06/2021	09/06/2021
									A070504	Premature Discharge of Battery	09/06/2021	09/06/2021
							F1903	Device Explantation	A0705	Battery Problem	09/06/2021	09/06/2021
									A070504	Premature Discharge of Battery	09/06/2021	09/06/2021
							F1905	Device Revision or Replacement	A0705	Battery Problem	09/06/2021	09/06/2021
									A070504	Premature Discharge of Battery	09/06/2021	09/06/2021
					CT2727	Active implant devices/ancillaries	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem	09/06/2021	09/06/2021
									A070504	Premature Discharge of Battery	09/06/2021	09/06/2021
							F08	Hospitalization or Prolonged Hospitalization	A0705	Battery Problem	09/06/2021	09/06/2021
									A070504	Premature Discharge of Battery	09/06/2021	09/06/2021
							F12	Serious Injury/ Illness/ Impairment	A0705	Battery Problem	09/06/2021	09/06/2021
									A070504	Premature Discharge of Battery	09/06/2021	09/06/2021
							F1903	Device Explantation	A0705	Battery Problem	09/06/2021	09/06/2021
									A070504	Premature Discharge of Battery	09/06/2021	09/06/2021
							F1905	Device Revision or Replacement	A0705	Battery Problem	09/06/2021	09/06/2021
									A070504	Premature Discharge of Battery	09/06/2021	09/06/2021
					CT2750	Implantable cardiac electrical stimulators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem	09/06/2021	09/06/2021
									A070504	Premature Discharge of Battery	09/06/2021	09/06/2021
							F08	Hospitalization or Prolonged Hospitalization	A0705	Battery Problem	09/06/2021	09/06/2021
									A070504	Premature Discharge of Battery	09/06/2021	09/06/2021
							F12	Serious Injury/ Illness/ Impairment	A0705	Battery Problem	09/06/2021	09/06/2021
									A070504	Premature Discharge of Battery	09/06/2021	09/06/2021
							F1903	Device Explantation	A0705	Battery Problem	09/06/2021	09/06/2021
									A070504	Premature Discharge of Battery	09/06/2021	09/06/2021
							F1905	Device Revision or Replacement	A0705	Battery Problem	09/06/2021	09/06/2021
									A070504	Premature Discharge of Battery	09/06/2021	09/06/2021
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem	09/06/2021	09/06/2021
									A070504	Premature Discharge of Battery	09/06/2021	09/06/2021
							F08	Hospitalization or Prolonged Hospitalization	A0705	Battery Problem	09/06/2021	09/06/2021
									A070504	Premature Discharge of Battery	09/06/2021	09/06/2021
							F12	Serious Injury/ Illness/ Impairment	A0705	Battery Problem	09/06/2021	09/06/2021
									A070504	Premature Discharge of Battery	09/06/2021	09/06/2021
							F1903	Device Explantation	A0705	Battery Problem	09/06/2021	09/06/2021
									A070504	Premature Discharge of Battery	09/06/2021	09/06/2021
							F1905	Device Revision or Replacement	A0705	Battery Problem	09/06/2021	09/06/2021
									A070504	Premature Discharge of Battery	09/06/2021	09/06/2021
					CT2985	Active	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem	09/06/2021	09/06/2021
									A070504	Premature Discharge of Battery	09/06/2021	09/06/2021
							F08	Hospitalization or Prolonged Hospitalization	A0705	Battery Problem	09/06/2021	09/06/2021
									A070504	Premature Discharge of Battery	09/06/2021	09/06/2021
							F12	Serious Injury/ Illness/ Impairment	A0705	Battery Problem	09/06/2021	09/06/2021
									A070504	Premature Discharge of Battery	09/06/2021	09/06/2021
							F1903	Device Explantation	A0705	Battery Problem	09/06/2021	09/06/2021
									A070504	Premature Discharge of Battery	09/06/2021	09/06/2021
							F1905	Device Revision or Replacement	A0705	Battery Problem	09/06/2021	09/06/2021

									A070504	Premature Discharge of Battery	09/06/2021	09/06/2021
					CT3001	Implantable metal	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem	09/06/2021	09/06/2021
									A070504	Premature Discharge of Battery	09/06/2021	09/06/2021
							F08	Hospitalization or Prolonged Hospitalization	A0705	Battery Problem	09/06/2021	09/06/2021
									A070504	Premature Discharge of Battery	09/06/2021	09/06/2021
							F12	Serious Injury/ Illness/ Impairment	A0705	Battery Problem	09/06/2021	09/06/2021
									A070504	Premature Discharge of Battery	09/06/2021	09/06/2021
							F1903	Device Explantation	A0705	Battery Problem	09/06/2021	09/06/2021
									A070504	Premature Discharge of Battery	09/06/2021	09/06/2021
							F1905	Device Revision or Replacement	A0705	Battery Problem	09/06/2021	09/06/2021
									A070504	Premature Discharge of Battery	09/06/2021	09/06/2021
					CT3002	Implantable synthetic polymer	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem	09/06/2021	09/06/2021
									A070504	Premature Discharge of Battery	09/06/2021	09/06/2021
							F08	Hospitalization or Prolonged Hospitalization	A0705	Battery Problem	09/06/2021	09/06/2021
									A070504	Premature Discharge of Battery	09/06/2021	09/06/2021
							F12	Serious Injury/ Illness/ Impairment	A0705	Battery Problem	09/06/2021	09/06/2021
									A070504	Premature Discharge of Battery	09/06/2021	09/06/2021
							F1903	Device Explantation	A0705	Battery Problem	09/06/2021	09/06/2021
									A070504	Premature Discharge of Battery	09/06/2021	09/06/2021
							F1905	Device Revision or Replacement	A0705	Battery Problem	09/06/2021	09/06/2021
									A070504	Premature Discharge of Battery	09/06/2021	09/06/2021
					CT3006	Implantable inorganic material	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem	09/06/2021	09/06/2021
									A070504	Premature Discharge of Battery	09/06/2021	09/06/2021
							F08	Hospitalization or Prolonged Hospitalization	A0705	Battery Problem	09/06/2021	09/06/2021
									A070504	Premature Discharge of Battery	09/06/2021	09/06/2021
							F12	Serious Injury/ Illness/ Impairment	A0705	Battery Problem	09/06/2021	09/06/2021
									A070504	Premature Discharge of Battery	09/06/2021	09/06/2021
							F1903	Device Explantation	A0705	Battery Problem	09/06/2021	09/06/2021
									A070504	Premature Discharge of Battery	09/06/2021	09/06/2021
							F1905	Device Revision or Replacement	A0705	Battery Problem	09/06/2021	09/06/2021
									A070504	Premature Discharge of Battery	09/06/2021	09/06/2021
					CT314	Electrically powered	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem	09/06/2021	09/06/2021
									A070504	Premature Discharge of Battery	09/06/2021	09/06/2021
							F08	Hospitalization or Prolonged Hospitalization	A0705	Battery Problem	09/06/2021	09/06/2021
									A070504	Premature Discharge of Battery	09/06/2021	09/06/2021
							F12	Serious Injury/ Illness/ Impairment	A0705	Battery Problem	09/06/2021	09/06/2021
									A070504	Premature Discharge of Battery	09/06/2021	09/06/2021
							F1903	Device Explantation	A0705	Battery Problem	09/06/2021	09/06/2021
									A070504	Premature Discharge of Battery	09/06/2021	09/06/2021
							F1905	Device Revision or Replacement	A0705	Battery Problem	09/06/2021	09/06/2021
									A070504	Premature Discharge of Battery	09/06/2021	09/06/2021
					CT321	Long-term surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem	09/06/2021	09/06/2021
									A070504	Premature Discharge of Battery	09/06/2021	09/06/2021
							F08	Hospitalization or Prolonged Hospitalization	A0705	Battery Problem	09/06/2021	09/06/2021
									A070504	Premature Discharge of Battery	09/06/2021	09/06/2021
							F12	Serious Injury/ Illness/ Impairment	A0705	Battery Problem	09/06/2021	09/06/2021
									A070504	Premature Discharge of Battery	09/06/2021	09/06/2021
							F1903	Device Explantation	A0705	Battery Problem	09/06/2021	09/06/2021
									A070504	Premature Discharge of Battery	09/06/2021	09/06/2021
							F1905	Device Revision or Replacement	A0705	Battery Problem	09/06/2021	09/06/2021
									A070504	Premature Discharge of Battery	09/06/2021	09/06/2021
					CT449	Generators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem	09/06/2021	09/06/2021
									A070504	Premature Discharge of Battery	09/06/2021	09/06/2021
							F08	Hospitalization or Prolonged Hospitalization	A0705	Battery Problem	09/06/2021	09/06/2021
									A070504	Premature Discharge of Battery	09/06/2021	09/06/2021
							F12	Serious Injury/ Illness/ Impairment	A0705	Battery Problem	09/06/2021	09/06/2021
									A070504	Premature Discharge of Battery	09/06/2021	09/06/2021
							F1903	Device Explantation	A0705	Battery Problem	09/06/2021	09/06/2021
									A070504	Premature Discharge of Battery	09/06/2021	09/06/2021
							F1905	Device Revision or Replacement	A0705	Battery Problem	09/06/2021	09/06/2021
									A070504	Premature Discharge of Battery	09/06/2021	09/06/2021
					CT981	Single use	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem	09/06/2021	09/06/2021
									A070504	Premature Discharge of Battery	09/06/2021	09/06/2021
							F08	Hospitalization or Prolonged Hospitalization	A0705	Battery Problem	09/06/2021	09/06/2021
									A070504	Premature Discharge of Battery	09/06/2021	09/06/2021
							F12	Serious Injury/ Illness/ Impairment	A0705	Battery Problem	09/06/2021	09/06/2021
									A070504	Premature Discharge of Battery	09/06/2021	09/06/2021
							F1903	Device Explantation	A0705	Battery Problem	09/06/2021	09/06/2021
									A070504	Premature Discharge of Battery	09/06/2021	09/06/2021
							F1905	Device Revision or Replacement	A0705	Battery Problem	09/06/2021	09/06/2021
									A070504	Premature Discharge of Battery	09/06/2021	09/06/2021
					CT983	Surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem	09/06/2021	09/06/2021
									A070504	Premature Discharge of Battery	09/06/2021	09/06/2021
							F08	Hospitalization or Prolonged Hospitalization	A0705	Battery Problem	09/06/2021	09/06/2021
									A070504	Premature Discharge of Battery	09/06/2021	09/06/2021
							F12	Serious Injury/ Illness/ Impairment	A0705	Battery Problem	09/06/2021	09/06/2021
									A070504	Premature Discharge of Battery	09/06/2021	09/06/2021
							F1903	Device Explantation	A0705	Battery Problem	09/06/2021	09/06/2021
									A070504	Premature Discharge of Battery	09/06/2021	09/06/2021
							F1905	Device Revision or Replacement	A0705	Battery Problem	09/06/2021	09/06/2021
									A070504	Premature Discharge of Battery	09/06/2021	09/06/2021
11253	2021	All other reportable incidents	47270	Cardiac resynchronization therapy implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	14/06/2021	14/06/2021
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery	14/06/2021	14/06/2021
11254	2021	All other reportable incidents	37265	Dual-chamber implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	23/06/2021	23/06/2021
							F1901	Additional Surgery	A070504	Premature Discharge of Battery	23/06/2021	23/06/2021
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	23/06/2021	23/06/2021
11255	2021	All other reportable incidents	35852	Single-chamber implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	27/04/2021	27/04/2021
									A160503	Inappropriate or Unexpected Reset	27/04/2021	27/04/2021

							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery	27/04/2021	27/04/2021
									A160503	Inappropriate or Unexpected Reset	27/04/2021	27/04/2021
11256	2021	All other reportable incidents	36241	Pacemaker/defibrillator lead	CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A040101	Fracture	28/05/2021	28/05/2021
									A070909	Oversensing	28/05/2021	28/05/2021
									A0712	Pacing Problem	28/05/2021	28/05/2021
							F08	Hospitalization or Prolonged Hospitalization	A040101	Fracture	28/05/2021	28/05/2021
									A070909	Oversensing	28/05/2021	28/05/2021
									A0712	Pacing Problem	28/05/2021	28/05/2021
							F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A040101	Fracture	28/05/2021	28/05/2021
									A070909	Oversensing	28/05/2021	28/05/2021
									A0712	Pacing Problem	28/05/2021	28/05/2021
							F1905	Device Revision or Replacement	A040101	Fracture	28/05/2021	28/05/2021
									A070909	Oversensing	28/05/2021	28/05/2021
									A0712	Pacing Problem	28/05/2021	28/05/2021
11257	2021	All other reportable incidents	35852	Single-chamber implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem	21/06/2021	06/01/2022
									A070504	Premature Discharge of Battery	21/06/2021	06/01/2022
							F08	Hospitalization or Prolonged Hospitalization	A0705	Battery Problem	21/06/2021	06/01/2022
									A070504	Premature Discharge of Battery	21/06/2021	06/01/2022
							F1905	Device Revision or Replacement	A0705	Battery Problem	21/06/2021	06/01/2022
									A070504	Premature Discharge of Battery	21/06/2021	06/01/2022
11258	2021	All other reportable incidents	47265	Dual-chamber implantable pacemaker, rate-responsive	CT141	Electrodes/Leads and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
					CT145	Cardiovascular devices	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
					CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
					CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
					CT233	Surgical	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
					CT2406	Implantable	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
					CT2407	Active implantable	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
					CT2727	Active implant devices/ancillaries	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
					CT2750	Implantable cardiac electrical stimulators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
					CT2985	Active	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
					CT3001	Implantable metal	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
					CT3002	Implantable synthetic polymer	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
					CT3006	Implantable inorganic material	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
					CT314	Electrically powered	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
					CT321	Long-term surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		

					CT449	Generators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
					CT981	Single use	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
					CT983	Surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
11259	2021	All other reportable incidents	47267	Single-chamber implantable pacemaker, rate-responsive	CT141	Electrodes/Leads and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem	10/11/2021	10/11/2021
									A070504	Premature Discharge of Battery	10/11/2021	10/11/2021
							F08	Hospitalization or Prolonged Hospitalization	A0705	Battery Problem	10/11/2021	10/11/2021
									A070504	Premature Discharge of Battery	10/11/2021	10/11/2021
							F1905	Device Revision or Replacement	A0705	Battery Problem	10/11/2021	10/11/2021
									A070504	Premature Discharge of Battery	10/11/2021	10/11/2021
					CT145	Cardiovascular devices	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem	10/11/2021	10/11/2021
									A070504	Premature Discharge of Battery	10/11/2021	10/11/2021
							F08	Hospitalization or Prolonged Hospitalization	A0705	Battery Problem	10/11/2021	10/11/2021
									A070504	Premature Discharge of Battery	10/11/2021	10/11/2021
							F1905	Device Revision or Replacement	A0705	Battery Problem	10/11/2021	10/11/2021
									A070504	Premature Discharge of Battery	10/11/2021	10/11/2021
					CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem	10/11/2021	10/11/2021
									A070504	Premature Discharge of Battery	10/11/2021	10/11/2021
							F08	Hospitalization or Prolonged Hospitalization	A0705	Battery Problem	10/11/2021	10/11/2021
									A070504	Premature Discharge of Battery	10/11/2021	10/11/2021
							F1905	Device Revision or Replacement	A0705	Battery Problem	10/11/2021	10/11/2021
									A070504	Premature Discharge of Battery	10/11/2021	10/11/2021
					CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem	10/11/2021	10/11/2021
									A070504	Premature Discharge of Battery	10/11/2021	10/11/2021
							F08	Hospitalization or Prolonged Hospitalization	A0705	Battery Problem	10/11/2021	10/11/2021
									A070504	Premature Discharge of Battery	10/11/2021	10/11/2021
							F1905	Device Revision or Replacement	A0705	Battery Problem	10/11/2021	10/11/2021
									A070504	Premature Discharge of Battery	10/11/2021	10/11/2021
					CT233	Surgical	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem	10/11/2021	10/11/2021
									A070504	Premature Discharge of Battery	10/11/2021	10/11/2021
							F08	Hospitalization or Prolonged Hospitalization	A0705	Battery Problem	10/11/2021	10/11/2021
									A070504	Premature Discharge of Battery	10/11/2021	10/11/2021
							F1905	Device Revision or Replacement	A0705	Battery Problem	10/11/2021	10/11/2021
									A070504	Premature Discharge of Battery	10/11/2021	10/11/2021
					CT2406	Implantable	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem	10/11/2021	10/11/2021
									A070504	Premature Discharge of Battery	10/11/2021	10/11/2021
							F08	Hospitalization or Prolonged Hospitalization	A0705	Battery Problem	10/11/2021	10/11/2021
									A070504	Premature Discharge of Battery	10/11/2021	10/11/2021
							F1905	Device Revision or Replacement	A0705	Battery Problem	10/11/2021	10/11/2021
									A070504	Premature Discharge of Battery	10/11/2021	10/11/2021
					CT2407	Active implantable	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem	10/11/2021	10/11/2021
									A070504	Premature Discharge of Battery	10/11/2021	10/11/2021
							F08	Hospitalization or Prolonged Hospitalization	A0705	Battery Problem	10/11/2021	10/11/2021
									A070504	Premature Discharge of Battery	10/11/2021	10/11/2021
							F1905	Device Revision or Replacement	A0705	Battery Problem	10/11/2021	10/11/2021
									A070504	Premature Discharge of Battery	10/11/2021	10/11/2021
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem	10/11/2021	10/11/2021
									A070504	Premature Discharge of Battery	10/11/2021	10/11/2021
							F08	Hospitalization or Prolonged Hospitalization	A0705	Battery Problem	10/11/2021	10/11/2021
									A070504	Premature Discharge of Battery	10/11/2021	10/11/2021
							F1905	Device Revision or Replacement	A0705	Battery Problem	10/11/2021	10/11/2021
									A070504	Premature Discharge of Battery	10/11/2021	10/11/2021
					CT2727	Active implant devices/ancillaries	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem	10/11/2021	10/11/2021
									A070504	Premature Discharge of Battery	10/11/2021	10/11/2021
							F08	Hospitalization or Prolonged Hospitalization	A0705	Battery Problem	10/11/2021	10/11/2021
									A070504	Premature Discharge of Battery	10/11/2021	10/11/2021
							F1905	Device Revision or Replacement	A0705	Battery Problem	10/11/2021	10/11/2021
									A070504	Premature Discharge of Battery	10/11/2021	10/11/2021
					CT2750	Implantable cardiac electrical stimulators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem	10/11/2021	10/11/2021
									A070504	Premature Discharge of Battery	10/11/2021	10/11/2021
							F08	Hospitalization or Prolonged Hospitalization	A0705	Battery Problem	10/11/2021	10/11/2021
									A070504	Premature Discharge of Battery	10/11/2021	10/11/2021
							F1905	Device Revision or Replacement	A0705	Battery Problem	10/11/2021	10/11/2021
									A070504	Premature Discharge of Battery	10/11/2021	10/11/2021
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem	10/11/2021	10/11/2021
									A070504	Premature Discharge of Battery	10/11/2021	10/11/2021
							F08	Hospitalization or Prolonged Hospitalization	A0705	Battery Problem	10/11/2021	10/11/2021
									A070504	Premature Discharge of Battery	10/11/2021	10/11/2021
							F1905	Device Revision or Replacement	A0705	Battery Problem	10/11/2021	10/11/2021
									A070504	Premature Discharge of Battery	10/11/2021	10/11/2021
					CT2985	Active	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem	10/11/2021	10/11/2021
									A070504	Premature Discharge of Battery	10/11/2021	10/11/2021
							F08	Hospitalization or Prolonged Hospitalization	A0705	Battery Problem	10/11/2021	10/11/2021
									A070504	Premature Discharge of Battery	10/11/2021	10/11/2021
							F1905	Device Revision or Replacement	A0705	Battery Problem	10/11/2021	10/11/2021
									A070504	Premature Discharge of Battery	10/11/2021	10/11/2021
					CT3001	Implantable metal	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem	10/11/2021	10/11/2021
									A070504	Premature Discharge of Battery	10/11/2021	10/11/2021
							F08	Hospitalization or Prolonged Hospitalization	A0705	Battery Problem	10/11/2021	10/11/2021
									A070504	Premature Discharge of Battery	10/11/2021	10/11/2021

							F1905	Device Revision or Replacement	A0705	Battery Problem	10/11/2021	10/11/2021
									A070504	Premature Discharge of Battery	10/11/2021	10/11/2021
					CT3002	Implantable synthetic polymer	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem	10/11/2021	10/11/2021
									A070504	Premature Discharge of Battery	10/11/2021	10/11/2021
							F08	Hospitalization or Prolonged Hospitalization	A0705	Battery Problem	10/11/2021	10/11/2021
									A070504	Premature Discharge of Battery	10/11/2021	10/11/2021
							F1905	Device Revision or Replacement	A0705	Battery Problem	10/11/2021	10/11/2021
									A070504	Premature Discharge of Battery	10/11/2021	10/11/2021
					CT3006	Implantable inorganic material	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem	10/11/2021	10/11/2021
									A070504	Premature Discharge of Battery	10/11/2021	10/11/2021
							F08	Hospitalization or Prolonged Hospitalization	A0705	Battery Problem	10/11/2021	10/11/2021
									A070504	Premature Discharge of Battery	10/11/2021	10/11/2021
							F1905	Device Revision or Replacement	A0705	Battery Problem	10/11/2021	10/11/2021
									A070504	Premature Discharge of Battery	10/11/2021	10/11/2021
					CT314	Electrically powered	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem	10/11/2021	10/11/2021
									A070504	Premature Discharge of Battery	10/11/2021	10/11/2021
							F08	Hospitalization or Prolonged Hospitalization	A0705	Battery Problem	10/11/2021	10/11/2021
									A070504	Premature Discharge of Battery	10/11/2021	10/11/2021
							F1905	Device Revision or Replacement	A0705	Battery Problem	10/11/2021	10/11/2021
									A070504	Premature Discharge of Battery	10/11/2021	10/11/2021
					CT321	Long-term surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem	10/11/2021	10/11/2021
									A070504	Premature Discharge of Battery	10/11/2021	10/11/2021
							F08	Hospitalization or Prolonged Hospitalization	A0705	Battery Problem	10/11/2021	10/11/2021
									A070504	Premature Discharge of Battery	10/11/2021	10/11/2021
							F1905	Device Revision or Replacement	A0705	Battery Problem	10/11/2021	10/11/2021
									A070504	Premature Discharge of Battery	10/11/2021	10/11/2021
					CT449	Generators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem	10/11/2021	10/11/2021
									A070504	Premature Discharge of Battery	10/11/2021	10/11/2021
							F08	Hospitalization or Prolonged Hospitalization	A0705	Battery Problem	10/11/2021	10/11/2021
									A070504	Premature Discharge of Battery	10/11/2021	10/11/2021
							F1905	Device Revision or Replacement	A0705	Battery Problem	10/11/2021	10/11/2021
									A070504	Premature Discharge of Battery	10/11/2021	10/11/2021
					CT981	Single use	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem	10/11/2021	10/11/2021
									A070504	Premature Discharge of Battery	10/11/2021	10/11/2021
							F08	Hospitalization or Prolonged Hospitalization	A0705	Battery Problem	10/11/2021	10/11/2021
									A070504	Premature Discharge of Battery	10/11/2021	10/11/2021
							F1905	Device Revision or Replacement	A0705	Battery Problem	10/11/2021	10/11/2021
									A070504	Premature Discharge of Battery	10/11/2021	10/11/2021
					CT983	Surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem	10/11/2021	10/11/2021
									A070504	Premature Discharge of Battery	10/11/2021	10/11/2021
							F08	Hospitalization or Prolonged Hospitalization	A0705	Battery Problem	10/11/2021	10/11/2021
									A070504	Premature Discharge of Battery	10/11/2021	10/11/2021
							F1905	Device Revision or Replacement	A0705	Battery Problem	10/11/2021	10/11/2021
									A070504	Premature Discharge of Battery	10/11/2021	10/11/2021
11260	2021	All other reportable incidents	47265	Dual-chamber implantable pacemaker, rate-responsive	CT141	Electrodes/Leads and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem	24/06/2021	24/06/2021
									A070504	Premature Discharge of Battery	24/06/2021	24/06/2021
							F08	Hospitalization or Prolonged Hospitalization	A0705	Battery Problem	24/06/2021	24/06/2021
									A070504	Premature Discharge of Battery	24/06/2021	24/06/2021
							F1905	Device Revision or Replacement	A0705	Battery Problem	24/06/2021	24/06/2021
									A070504	Premature Discharge of Battery	24/06/2021	24/06/2021
					CT145	Cardiovascular devices	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem	24/06/2021	24/06/2021
									A070504	Premature Discharge of Battery	24/06/2021	24/06/2021
							F08	Hospitalization or Prolonged Hospitalization	A0705	Battery Problem	24/06/2021	24/06/2021
									A070504	Premature Discharge of Battery	24/06/2021	24/06/2021
							F1905	Device Revision or Replacement	A0705	Battery Problem	24/06/2021	24/06/2021
									A070504	Premature Discharge of Battery	24/06/2021	24/06/2021
					CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem	24/06/2021	24/06/2021
									A070504	Premature Discharge of Battery	24/06/2021	24/06/2021
							F08	Hospitalization or Prolonged Hospitalization	A0705	Battery Problem	24/06/2021	24/06/2021
									A070504	Premature Discharge of Battery	24/06/2021	24/06/2021
							F1905	Device Revision or Replacement	A0705	Battery Problem	24/06/2021	24/06/2021
									A070504	Premature Discharge of Battery	24/06/2021	24/06/2021
					CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem	24/06/2021	24/06/2021
									A070504	Premature Discharge of Battery	24/06/2021	24/06/2021
							F08	Hospitalization or Prolonged Hospitalization	A0705	Battery Problem	24/06/2021	24/06/2021
									A070504	Premature Discharge of Battery	24/06/2021	24/06/2021
							F1905	Device Revision or Replacement	A0705	Battery Problem	24/06/2021	24/06/2021
									A070504	Premature Discharge of Battery	24/06/2021	24/06/2021
					CT233	Surgical	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem	24/06/2021	24/06/2021
									A070504	Premature Discharge of Battery	24/06/2021	24/06/2021
							F08	Hospitalization or Prolonged Hospitalization	A0705	Battery Problem	24/06/2021	24/06/2021
									A070504	Premature Discharge of Battery	24/06/2021	24/06/2021
							F1905	Device Revision or Replacement	A0705	Battery Problem	24/06/2021	24/06/2021
									A070504	Premature Discharge of Battery	24/06/2021	24/06/2021
					CT2406	Implantable	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem	24/06/2021	24/06/2021
									A070504	Premature Discharge of Battery	24/06/2021	24/06/2021
							F08	Hospitalization or Prolonged Hospitalization	A0705	Battery Problem	24/06/2021	24/06/2021
									A070504	Premature Discharge of Battery	24/06/2021	24/06/2021
							F1905	Device Revision or Replacement	A0705	Battery Problem	24/06/2021	24/06/2021
									A070504	Premature Discharge of Battery	24/06/2021	24/06/2021
					CT2407	Active implantable	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem	24/06/2021	24/06/2021
									A070504	Premature Discharge of Battery	24/06/2021	24/06/2021
							F08	Hospitalization or Prolonged Hospitalization	A0705	Battery Problem	24/06/2021	24/06/2021
									A070504	Premature Discharge of Battery	24/06/2021	24/06/2021
							F1905	Device Revision or Replacement	A0705	Battery Problem	24/06/2021	24/06/2021
									A070504	Premature Discharge of Battery	24/06/2021	24/06/2021
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem	24/06/2021	24/06/2021
									A070504	Premature Discharge of Battery	24/06/2021	24/06/2021

							F08	Hospitalization or Prolonged Hospitalization	A0705	Battery Problem	24/06/2021	24/06/2021
									A070504	Premature Discharge of Battery	24/06/2021	24/06/2021
							F1905	Device Revision or Replacement	A0705	Battery Problem	24/06/2021	24/06/2021
									A070504	Premature Discharge of Battery	24/06/2021	24/06/2021
					CT2727	Active implant devices/ancillaries	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem	24/06/2021	24/06/2021
									A070504	Premature Discharge of Battery	24/06/2021	24/06/2021
							F08	Hospitalization or Prolonged Hospitalization	A0705	Battery Problem	24/06/2021	24/06/2021
									A070504	Premature Discharge of Battery	24/06/2021	24/06/2021
							F1905	Device Revision or Replacement	A0705	Battery Problem	24/06/2021	24/06/2021
									A070504	Premature Discharge of Battery	24/06/2021	24/06/2021
					CT2750	Implantable cardiac electrical stimulators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem	24/06/2021	24/06/2021
									A070504	Premature Discharge of Battery	24/06/2021	24/06/2021
							F08	Hospitalization or Prolonged Hospitalization	A0705	Battery Problem	24/06/2021	24/06/2021
									A070504	Premature Discharge of Battery	24/06/2021	24/06/2021
							F1905	Device Revision or Replacement	A0705	Battery Problem	24/06/2021	24/06/2021
									A070504	Premature Discharge of Battery	24/06/2021	24/06/2021
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem	24/06/2021	24/06/2021
									A070504	Premature Discharge of Battery	24/06/2021	24/06/2021
							F08	Hospitalization or Prolonged Hospitalization	A0705	Battery Problem	24/06/2021	24/06/2021
									A070504	Premature Discharge of Battery	24/06/2021	24/06/2021
							F1905	Device Revision or Replacement	A0705	Battery Problem	24/06/2021	24/06/2021
									A070504	Premature Discharge of Battery	24/06/2021	24/06/2021
					CT2985	Active	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem	24/06/2021	24/06/2021
									A070504	Premature Discharge of Battery	24/06/2021	24/06/2021
							F08	Hospitalization or Prolonged Hospitalization	A0705	Battery Problem	24/06/2021	24/06/2021
									A070504	Premature Discharge of Battery	24/06/2021	24/06/2021
							F1905	Device Revision or Replacement	A0705	Battery Problem	24/06/2021	24/06/2021
									A070504	Premature Discharge of Battery	24/06/2021	24/06/2021
					CT3001	Implantable metal	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem	24/06/2021	24/06/2021
									A070504	Premature Discharge of Battery	24/06/2021	24/06/2021
							F08	Hospitalization or Prolonged Hospitalization	A0705	Battery Problem	24/06/2021	24/06/2021
									A070504	Premature Discharge of Battery	24/06/2021	24/06/2021
							F1905	Device Revision or Replacement	A0705	Battery Problem	24/06/2021	24/06/2021
									A070504	Premature Discharge of Battery	24/06/2021	24/06/2021
					CT3002	Implantable synthetic polymer	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem	24/06/2021	24/06/2021
									A070504	Premature Discharge of Battery	24/06/2021	24/06/2021
							F08	Hospitalization or Prolonged Hospitalization	A0705	Battery Problem	24/06/2021	24/06/2021
									A070504	Premature Discharge of Battery	24/06/2021	24/06/2021
							F1905	Device Revision or Replacement	A0705	Battery Problem	24/06/2021	24/06/2021
									A070504	Premature Discharge of Battery	24/06/2021	24/06/2021
					CT3006	Implantable inorganic material	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem	24/06/2021	24/06/2021
									A070504	Premature Discharge of Battery	24/06/2021	24/06/2021
							F08	Hospitalization or Prolonged Hospitalization	A0705	Battery Problem	24/06/2021	24/06/2021
									A070504	Premature Discharge of Battery	24/06/2021	24/06/2021
							F1905	Device Revision or Replacement	A0705	Battery Problem	24/06/2021	24/06/2021
									A070504	Premature Discharge of Battery	24/06/2021	24/06/2021
					CT314	Electrically powered	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem	24/06/2021	24/06/2021
									A070504	Premature Discharge of Battery	24/06/2021	24/06/2021
							F08	Hospitalization or Prolonged Hospitalization	A0705	Battery Problem	24/06/2021	24/06/2021
									A070504	Premature Discharge of Battery	24/06/2021	24/06/2021
							F1905	Device Revision or Replacement	A0705	Battery Problem	24/06/2021	24/06/2021
									A070504	Premature Discharge of Battery	24/06/2021	24/06/2021
					CT321	Long-term surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem	24/06/2021	24/06/2021
									A070504	Premature Discharge of Battery	24/06/2021	24/06/2021
							F08	Hospitalization or Prolonged Hospitalization	A0705	Battery Problem	24/06/2021	24/06/2021
									A070504	Premature Discharge of Battery	24/06/2021	24/06/2021
							F1905	Device Revision or Replacement	A0705	Battery Problem	24/06/2021	24/06/2021
									A070504	Premature Discharge of Battery	24/06/2021	24/06/2021
					CT449	Generators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem	24/06/2021	24/06/2021
									A070504	Premature Discharge of Battery	24/06/2021	24/06/2021
							F08	Hospitalization or Prolonged Hospitalization	A0705	Battery Problem	24/06/2021	24/06/2021
									A070504	Premature Discharge of Battery	24/06/2021	24/06/2021
							F1905	Device Revision or Replacement	A0705	Battery Problem	24/06/2021	24/06/2021
									A070504	Premature Discharge of Battery	24/06/2021	24/06/2021
					CT981	Single use	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem	24/06/2021	24/06/2021
									A070504	Premature Discharge of Battery	24/06/2021	24/06/2021
							F08	Hospitalization or Prolonged Hospitalization	A0705	Battery Problem	24/06/2021	24/06/2021
									A070504	Premature Discharge of Battery	24/06/2021	24/06/2021
							F1905	Device Revision or Replacement	A0705	Battery Problem	24/06/2021	24/06/2021
									A070504	Premature Discharge of Battery	24/06/2021	24/06/2021
					CT983	Surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem	24/06/2021	24/06/2021
									A070504	Premature Discharge of Battery	24/06/2021	24/06/2021
							F08	Hospitalization or Prolonged Hospitalization	A0705	Battery Problem	24/06/2021	24/06/2021
									A070504	Premature Discharge of Battery	24/06/2021	24/06/2021
							F1905	Device Revision or Replacement	A0705	Battery Problem	24/06/2021	24/06/2021
									A070504	Premature Discharge of Battery	24/06/2021	24/06/2021
11261	2021	All other reportable incidents	37265	Dual-chamber implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
11262	2021	All other reportable incidents	35852	Single-chamber implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem	15/09/2021	15/09/2021
									A070504	Premature Discharge of Battery	15/09/2021	15/09/2021
							F08	Hospitalization or Prolonged Hospitalization	A0705	Battery Problem	15/09/2021	15/09/2021
									A070504	Premature Discharge of Battery	15/09/2021	15/09/2021
							F1905	Device Revision or Replacement	A0705	Battery Problem	15/09/2021	15/09/2021
									A070504	Premature Discharge of Battery	15/09/2021	15/09/2021
11263	2021	All other reportable incidents	35852	Single-chamber implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	01/07/2021	01/07/2021
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	01/07/2021	01/07/2021

							F1903	Device Explantation	A070504	Premature Discharge of Battery	01/07/2021	01/07/2021
11264	2021	All other reportable incidents	47270	Cardiac resynchronization therapy implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E2104	Electric Shock	A071302	Inappropriate/Inadequate Shock/Stimulation		
							E2403	No Clinical Signs, Symptoms or Conditions	A071302	Inappropriate/Inadequate Shock/Stimulation		
							F26	No Health Consequences or Impact	A071302	Inappropriate/Inadequate Shock/Stimulation		
11265	2021	All other reportable incidents	47265	Dual-chamber implantable pacemaker, rate-responsive	CT141	Electrodes/Leads and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem	02/08/2021	02/08/2021
									A070504	Premature Discharge of Battery	02/08/2021	02/08/2021
							F08	Hospitalization or Prolonged Hospitalization	A0705	Battery Problem	02/08/2021	02/08/2021
									A070504	Premature Discharge of Battery	02/08/2021	02/08/2021
							F1905	Device Revision or Replacement	A0705	Battery Problem	02/08/2021	02/08/2021
									A070504	Premature Discharge of Battery	02/08/2021	02/08/2021
					CT145	Cardiovascular devices	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem	02/08/2021	02/08/2021
									A070504	Premature Discharge of Battery	02/08/2021	02/08/2021
							F08	Hospitalization or Prolonged Hospitalization	A0705	Battery Problem	02/08/2021	02/08/2021
									A070504	Premature Discharge of Battery	02/08/2021	02/08/2021
							F1905	Device Revision or Replacement	A0705	Battery Problem	02/08/2021	02/08/2021
									A070504	Premature Discharge of Battery	02/08/2021	02/08/2021
					CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem	02/08/2021	02/08/2021
									A070504	Premature Discharge of Battery	02/08/2021	02/08/2021
							F08	Hospitalization or Prolonged Hospitalization	A0705	Battery Problem	02/08/2021	02/08/2021
									A070504	Premature Discharge of Battery	02/08/2021	02/08/2021
							F1905	Device Revision or Replacement	A0705	Battery Problem	02/08/2021	02/08/2021
									A070504	Premature Discharge of Battery	02/08/2021	02/08/2021
					CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem	02/08/2021	02/08/2021
									A070504	Premature Discharge of Battery	02/08/2021	02/08/2021
							F08	Hospitalization or Prolonged Hospitalization	A0705	Battery Problem	02/08/2021	02/08/2021
									A070504	Premature Discharge of Battery	02/08/2021	02/08/2021
							F1905	Device Revision or Replacement	A0705	Battery Problem	02/08/2021	02/08/2021
									A070504	Premature Discharge of Battery	02/08/2021	02/08/2021
					CT233	Surgical	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem	02/08/2021	02/08/2021
									A070504	Premature Discharge of Battery	02/08/2021	02/08/2021
							F08	Hospitalization or Prolonged Hospitalization	A0705	Battery Problem	02/08/2021	02/08/2021
									A070504	Premature Discharge of Battery	02/08/2021	02/08/2021
							F1905	Device Revision or Replacement	A0705	Battery Problem	02/08/2021	02/08/2021
									A070504	Premature Discharge of Battery	02/08/2021	02/08/2021
					CT2406	Implantable	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem	02/08/2021	02/08/2021
									A070504	Premature Discharge of Battery	02/08/2021	02/08/2021
							F08	Hospitalization or Prolonged Hospitalization	A0705	Battery Problem	02/08/2021	02/08/2021
									A070504	Premature Discharge of Battery	02/08/2021	02/08/2021
							F1905	Device Revision or Replacement	A0705	Battery Problem	02/08/2021	02/08/2021
									A070504	Premature Discharge of Battery	02/08/2021	02/08/2021
					CT2407	Active implantable	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem	02/08/2021	02/08/2021
									A070504	Premature Discharge of Battery	02/08/2021	02/08/2021
							F08	Hospitalization or Prolonged Hospitalization	A0705	Battery Problem	02/08/2021	02/08/2021
									A070504	Premature Discharge of Battery	02/08/2021	02/08/2021
							F1905	Device Revision or Replacement	A0705	Battery Problem	02/08/2021	02/08/2021
									A070504	Premature Discharge of Battery	02/08/2021	02/08/2021
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem	02/08/2021	02/08/2021
									A070504	Premature Discharge of Battery	02/08/2021	02/08/2021
							F08	Hospitalization or Prolonged Hospitalization	A0705	Battery Problem	02/08/2021	02/08/2021
									A070504	Premature Discharge of Battery	02/08/2021	02/08/2021
							F1905	Device Revision or Replacement	A0705	Battery Problem	02/08/2021	02/08/2021
									A070504	Premature Discharge of Battery	02/08/2021	02/08/2021
					CT2727	Active implant devices/ancillaries	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem	02/08/2021	02/08/2021
									A070504	Premature Discharge of Battery	02/08/2021	02/08/2021
							F08	Hospitalization or Prolonged Hospitalization	A0705	Battery Problem	02/08/2021	02/08/2021
									A070504	Premature Discharge of Battery	02/08/2021	02/08/2021
							F1905	Device Revision or Replacement	A0705	Battery Problem	02/08/2021	02/08/2021
									A070504	Premature Discharge of Battery	02/08/2021	02/08/2021
					CT2750	Implantable cardiac electrical stimulators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem	02/08/2021	02/08/2021
									A070504	Premature Discharge of Battery	02/08/2021	02/08/2021
							F08	Hospitalization or Prolonged Hospitalization	A0705	Battery Problem	02/08/2021	02/08/2021
									A070504	Premature Discharge of Battery	02/08/2021	02/08/2021
							F1905	Device Revision or Replacement	A0705	Battery Problem	02/08/2021	02/08/2021
									A070504	Premature Discharge of Battery	02/08/2021	02/08/2021
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem	02/08/2021	02/08/2021
									A070504	Premature Discharge of Battery	02/08/2021	02/08/2021
							F08	Hospitalization or Prolonged Hospitalization	A0705	Battery Problem	02/08/2021	02/08/2021
									A070504	Premature Discharge of Battery	02/08/2021	02/08/2021
							F1905	Device Revision or Replacement	A0705	Battery Problem	02/08/2021	02/08/2021
									A070504	Premature Discharge of Battery	02/08/2021	02/08/2021
					CT2985	Active	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem	02/08/2021	02/08/2021
									A070504	Premature Discharge of Battery	02/08/2021	02/08/2021
							F08	Hospitalization or Prolonged Hospitalization	A0705	Battery Problem	02/08/2021	02/08/2021
									A070504	Premature Discharge of Battery	02/08/2021	02/08/2021
							F1905	Device Revision or Replacement	A0705	Battery Problem	02/08/2021	02/08/2021
									A070504	Premature Discharge of Battery	02/08/2021	02/08/2021
					CT3001	Implantable metal	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem	02/08/2021	02/08/2021
									A070504	Premature Discharge of Battery	02/08/2021	02/08/2021
							F08	Hospitalization or Prolonged Hospitalization	A0705	Battery Problem	02/08/2021	02/08/2021
									A070504	Premature Discharge of Battery	02/08/2021	02/08/2021
							F1905	Device Revision or Replacement	A0705	Battery Problem	02/08/2021	02/08/2021
									A070504	Premature Discharge of Battery	02/08/2021	02/08/2021
					CT3002	Implantable synthetic polymer	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem	02/08/2021	02/08/2021
									A070504	Premature Discharge of Battery	02/08/2021	02/08/2021
							F08	Hospitalization or Prolonged Hospitalization	A0705	Battery Problem	02/08/2021	02/08/2021
									A070504	Premature Discharge of Battery	02/08/2021	02/08/2021
							F1905	Device Revision or Replacement	A0705	Battery Problem	02/08/2021	02/08/2021
									A070504	Premature Discharge of Battery	02/08/2021	02/08/2021

					CT3006	Implantable inorganic material	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem	02/08/2021	02/08/2021
									A070504	Premature Discharge of Battery	02/08/2021	02/08/2021
							F08	Hospitalization or Prolonged Hospitalization	A0705	Battery Problem	02/08/2021	02/08/2021
									A070504	Premature Discharge of Battery	02/08/2021	02/08/2021
							F1905	Device Revision or Replacement	A0705	Battery Problem	02/08/2021	02/08/2021
									A070504	Premature Discharge of Battery	02/08/2021	02/08/2021
					CT314	Electrically powered	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem	02/08/2021	02/08/2021
									A070504	Premature Discharge of Battery	02/08/2021	02/08/2021
							F08	Hospitalization or Prolonged Hospitalization	A0705	Battery Problem	02/08/2021	02/08/2021
									A070504	Premature Discharge of Battery	02/08/2021	02/08/2021
							F1905	Device Revision or Replacement	A0705	Battery Problem	02/08/2021	02/08/2021
									A070504	Premature Discharge of Battery	02/08/2021	02/08/2021
					CT321	Long-term surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem	02/08/2021	02/08/2021
									A070504	Premature Discharge of Battery	02/08/2021	02/08/2021
							F08	Hospitalization or Prolonged Hospitalization	A0705	Battery Problem	02/08/2021	02/08/2021
									A070504	Premature Discharge of Battery	02/08/2021	02/08/2021
							F1905	Device Revision or Replacement	A0705	Battery Problem	02/08/2021	02/08/2021
									A070504	Premature Discharge of Battery	02/08/2021	02/08/2021
					CT449	Generators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem	02/08/2021	02/08/2021
									A070504	Premature Discharge of Battery	02/08/2021	02/08/2021
							F08	Hospitalization or Prolonged Hospitalization	A0705	Battery Problem	02/08/2021	02/08/2021
									A070504	Premature Discharge of Battery	02/08/2021	02/08/2021
							F1905	Device Revision or Replacement	A0705	Battery Problem	02/08/2021	02/08/2021
									A070504	Premature Discharge of Battery	02/08/2021	02/08/2021
					CT981	Single use	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem	02/08/2021	02/08/2021
									A070504	Premature Discharge of Battery	02/08/2021	02/08/2021
							F08	Hospitalization or Prolonged Hospitalization	A0705	Battery Problem	02/08/2021	02/08/2021
									A070504	Premature Discharge of Battery	02/08/2021	02/08/2021
							F1905	Device Revision or Replacement	A0705	Battery Problem	02/08/2021	02/08/2021
									A070504	Premature Discharge of Battery	02/08/2021	02/08/2021
					CT983	Surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem	02/08/2021	02/08/2021
									A070504	Premature Discharge of Battery	02/08/2021	02/08/2021
							F08	Hospitalization or Prolonged Hospitalization	A0705	Battery Problem	02/08/2021	02/08/2021
									A070504	Premature Discharge of Battery	02/08/2021	02/08/2021
							F1905	Device Revision or Replacement	A0705	Battery Problem	02/08/2021	02/08/2021
									A070504	Premature Discharge of Battery	02/08/2021	02/08/2021
11266	2021	All other reportable incidents	17882	Physiologic-monitoring defibrillation system	CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A07	Electrical /Electronic Property Problem		
							F27	Problem Identified During Non-Clinical Procedure	A07	Electrical /Electronic Property Problem		
11267	2021	All other reportable incidents	35852	Single-chamber implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
11268	2021	All other reportable incidents	35223	Endocardial/interventricular septal pacing lead	CT1684	Cardiac pacemakers and associated devices	E0604	Cardiac Perforation	A070102	High Capture Threshold		
									A0722	Impedance Problem		
									A072202	Low Impedance		
									A24	No Apparent Device Problem		
							E0717	Dyspnea	A070102	High Capture Threshold		
									A0722	Impedance Problem		
									A072202	Low Impedance		
									A24	No Apparent Device Problem		
							E0733	Pneumonia	A070102	High Capture Threshold		
									A0722	Impedance Problem		
									A072202	Low Impedance		
									A24	No Apparent Device Problem		
							E0752	Unspecified Respiratory Problem	A070102	High Capture Threshold		
									A0722	Impedance Problem		
									A072202	Low Impedance		
									A24	No Apparent Device Problem		
							E233001	Chest Pain	A070102	High Capture Threshold		
									A0722	Impedance Problem		
									A072202	Low Impedance		
									A24	No Apparent Device Problem		
							F11	Non-Serious Injury/ Illness/ Impairment	A070102	High Capture Threshold		
									A0722	Impedance Problem		
									A072202	Low Impedance		
									A24	No Apparent Device Problem		
							F1203	Life Threatening Illness or Injury	A070102	High Capture Threshold		
									A0722	Impedance Problem		
									A072202	Low Impedance		
									A24	No Apparent Device Problem		
11269	2021	All other reportable incidents	47265	Dual-chamber implantable pacemaker, rate-responsive	CT141	Electrodes/Leads and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	11/05/2021	11/05/2021
							F1901	Additional Surgery	A070504	Premature Discharge of Battery	11/05/2021	11/05/2021
					CT145	Cardiovascular devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	11/05/2021	11/05/2021
							F1901	Additional Surgery	A070504	Premature Discharge of Battery	11/05/2021	11/05/2021
					CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	11/05/2021	11/05/2021
							F1901	Additional Surgery	A070504	Premature Discharge of Battery	11/05/2021	11/05/2021
					CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	11/05/2021	11/05/2021
							F1901	Additional Surgery	A070504	Premature Discharge of Battery	11/05/2021	11/05/2021
					CT233	Surgical	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	11/05/2021	11/05/2021
							F1901	Additional Surgery	A070504	Premature Discharge of Battery	11/05/2021	11/05/2021
					CT2406	Implantable	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	11/05/2021	11/05/2021
							F1901	Additional Surgery	A070504	Premature Discharge of Battery	11/05/2021	11/05/2021
					CT2407	Active implantable	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	11/05/2021	11/05/2021
							F1901	Additional Surgery	A070504	Premature Discharge of Battery	11/05/2021	11/05/2021
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	11/05/2021	11/05/2021
							F1901	Additional Surgery	A070504	Premature Discharge of Battery	11/05/2021	11/05/2021
					CT2727	Active implant devices/ancillaries	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	11/05/2021	11/05/2021
							F1901	Additional Surgery	A070504	Premature Discharge of Battery	11/05/2021	11/05/2021
					CT2750	Implantable cardiac electrical stimulators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	11/05/2021	11/05/2021
							F1901	Additional Surgery	A070504	Premature Discharge of Battery	11/05/2021	11/05/2021

					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	11/05/2021	11/05/2021
							F1901	Additional Surgery	A070504	Premature Discharge of Battery	11/05/2021	11/05/2021
					CT2985	Active	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	11/05/2021	11/05/2021
							F1901	Additional Surgery	A070504	Premature Discharge of Battery	11/05/2021	11/05/2021
					CT3001	Implantable metal	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	11/05/2021	11/05/2021
							F1901	Additional Surgery	A070504	Premature Discharge of Battery	11/05/2021	11/05/2021
					CT3002	Implantable synthetic polymer	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	11/05/2021	11/05/2021
							F1901	Additional Surgery	A070504	Premature Discharge of Battery	11/05/2021	11/05/2021
					CT3006	Implantable inorganic material	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	11/05/2021	11/05/2021
							F1901	Additional Surgery	A070504	Premature Discharge of Battery	11/05/2021	11/05/2021
					CT314	Electrically powered	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	11/05/2021	11/05/2021
							F1901	Additional Surgery	A070504	Premature Discharge of Battery	11/05/2021	11/05/2021
					CT321	Long-term surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	11/05/2021	11/05/2021
							F1901	Additional Surgery	A070504	Premature Discharge of Battery	11/05/2021	11/05/2021
					CT449	Generators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	11/05/2021	11/05/2021
							F1901	Additional Surgery	A070504	Premature Discharge of Battery	11/05/2021	11/05/2021
					CT981	Single use	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	11/05/2021	11/05/2021
							F1901	Additional Surgery	A070504	Premature Discharge of Battery	11/05/2021	11/05/2021
					CT983	Surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	11/05/2021	11/05/2021
							F1901	Additional Surgery	A070504	Premature Discharge of Battery	11/05/2021	11/05/2021
11270	2021	All other reportable incidents	36241	Pacemaker/defibrillator lead	CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A040101	Fracture	17/07/2021	17/07/2021
									A070909	Oversensing	17/07/2021	17/07/2021
									A072201	High Impedance	17/07/2021	17/07/2021
									A090801	Signal Artifact/Noise	17/07/2021	17/07/2021
							F19	Surgical Intervention	A040101	Fracture	17/07/2021	17/07/2021
									A070909	Oversensing	17/07/2021	17/07/2021
									A072201	High Impedance	17/07/2021	17/07/2021
									A090801	Signal Artifact/Noise	17/07/2021	17/07/2021
							F1903	Device Explantation	A040101	Fracture	17/07/2021	17/07/2021
									A070909	Oversensing	17/07/2021	17/07/2021
									A072201	High Impedance	17/07/2021	17/07/2021
									A090801	Signal Artifact/Noise	17/07/2021	17/07/2021
11271	2021	All other reportable incidents	36241	Pacemaker/defibrillator lead	CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070909	Oversensing		
									A0712	Pacing Problem		
									A090801	Signal Artifact/Noise		
							F05	Delay to Treatment/ Therapy	A070909	Oversensing		
									A0712	Pacing Problem		
									A090801	Signal Artifact/Noise		
11272	2021	All other reportable incidents	35852	Single-chamber implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	18/05/2021	18/05/2021
							F1905	Device Revision or Replacement	A24	No Apparent Device Problem	18/05/2021	18/05/2021
							F26	No Health Consequences or Impact	A24	No Apparent Device Problem	18/05/2021	18/05/2021
11273	2021	All other reportable incidents	47263	Cardiac resynchronization therapy implantable pacemaker	CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem	18/06/2021	18/06/2021
									A070504	Premature Discharge of Battery	18/06/2021	18/06/2021
							F08	Hospitalization or Prolonged Hospitalization	A0705	Battery Problem	18/06/2021	18/06/2021
									A070504	Premature Discharge of Battery	18/06/2021	18/06/2021
							F1905	Device Revision or Replacement	A0705	Battery Problem	18/06/2021	18/06/2021
									A070504	Premature Discharge of Battery	18/06/2021	18/06/2021
					CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem	18/06/2021	18/06/2021
									A070504	Premature Discharge of Battery	18/06/2021	18/06/2021
							F08	Hospitalization or Prolonged Hospitalization	A0705	Battery Problem	18/06/2021	18/06/2021
									A070504	Premature Discharge of Battery	18/06/2021	18/06/2021
							F1905	Device Revision or Replacement	A0705	Battery Problem	18/06/2021	18/06/2021
									A070504	Premature Discharge of Battery	18/06/2021	18/06/2021
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem	18/06/2021	18/06/2021
									A070504	Premature Discharge of Battery	18/06/2021	18/06/2021
							F08	Hospitalization or Prolonged Hospitalization	A0705	Battery Problem	18/06/2021	18/06/2021
									A070504	Premature Discharge of Battery	18/06/2021	18/06/2021
							F1905	Device Revision or Replacement	A0705	Battery Problem	18/06/2021	18/06/2021
									A070504	Premature Discharge of Battery	18/06/2021	18/06/2021
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem	18/06/2021	18/06/2021
									A070504	Premature Discharge of Battery	18/06/2021	18/06/2021
							F08	Hospitalization or Prolonged Hospitalization	A0705	Battery Problem	18/06/2021	18/06/2021
									A070504	Premature Discharge of Battery	18/06/2021	18/06/2021
							F1905	Device Revision or Replacement	A0705	Battery Problem	18/06/2021	18/06/2021
									A070504	Premature Discharge of Battery	18/06/2021	18/06/2021
11274	2021	All other reportable incidents	17882	Physiologic-monitoring defibrillation system	CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A090801	Signal Artifact/Noise		
							F26	No Health Consequences or Impact	A090801	Signal Artifact/Noise		
11275	2021	All other reportable incidents	17882	Physiologic-monitoring defibrillation system	CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A07	Electrical /Electronic Property Problem		
							F27	Problem Identified During Non-Clinical Procedure	A07	Electrical /Electronic Property Problem		
11276	2021	All other reportable incidents	17882	Physiologic-monitoring defibrillation system	CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A1205	Loose or Intermittent Connection		
							F27	Problem Identified During Non-Clinical Procedure	A1205	Loose or Intermittent Connection		
11277	2021	All other reportable incidents	35852	Single-chamber implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A070909	Oversensing		
							F26	No Health Consequences or Impact	A070909	Oversensing		
11278	2021	All other reportable incidents	35852	Single-chamber implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	20/09/2021	20/09/2021
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	20/09/2021	20/09/2021
11279	2021	All other reportable incidents	37265	Dual-chamber implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A072201	High Impedance		
							F26	No Health Consequences or Impact	A072201	High Impedance		
11280	2021	All other reportable incidents	17882	Physiologic-monitoring defibrillation system	CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A1207	Incomplete or Inadequate Connection		
							F27	Problem Identified During Non-Clinical Procedure	A1207	Incomplete or Inadequate Connection		
11281	2021	All other reportable incidents	47265	Dual-chamber implantable pacemaker, rate-responsive	CT141	Electrodes/Leads and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	23/06/2021	21/09/2021
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	23/06/2021	21/09/2021
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	23/06/2021	21/09/2021
					CT145	Cardiovascular devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	23/06/2021	21/09/2021
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	23/06/2021	21/09/2021
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	23/06/2021	21/09/2021
					CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	23/06/2021	21/09/2021
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	23/06/2021	21/09/2021
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	23/06/2021	21/09/2021

					CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	23/06/2021	21/09/2021
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	23/06/2021	21/09/2021
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	23/06/2021	21/09/2021
					CT233	Surgical	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	23/06/2021	21/09/2021
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	23/06/2021	21/09/2021
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	23/06/2021	21/09/2021
					CT2406	Implantable	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	23/06/2021	21/09/2021
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	23/06/2021	21/09/2021
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	23/06/2021	21/09/2021
					CT2407	Active implantable	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	23/06/2021	21/09/2021
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	23/06/2021	21/09/2021
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	23/06/2021	21/09/2021
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	23/06/2021	21/09/2021
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	23/06/2021	21/09/2021
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	23/06/2021	21/09/2021
					CT2727	Active implant devices/ancillaries	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	23/06/2021	21/09/2021
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	23/06/2021	21/09/2021
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	23/06/2021	21/09/2021
					CT2750	Implantable cardiac electrical stimulators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	23/06/2021	21/09/2021
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	23/06/2021	21/09/2021
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	23/06/2021	21/09/2021
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	23/06/2021	21/09/2021
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	23/06/2021	21/09/2021
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	23/06/2021	21/09/2021
					CT2985	Active	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	23/06/2021	21/09/2021
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	23/06/2021	21/09/2021
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	23/06/2021	21/09/2021
					CT3001	Implantable metal	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	23/06/2021	21/09/2021
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	23/06/2021	21/09/2021
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	23/06/2021	21/09/2021
					CT3002	Implantable synthetic polymer	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	23/06/2021	21/09/2021
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	23/06/2021	21/09/2021
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	23/06/2021	21/09/2021
					CT3006	Implantable inorganic material	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	23/06/2021	21/09/2021
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	23/06/2021	21/09/2021
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	23/06/2021	21/09/2021
					CT314	Electrically powered	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	23/06/2021	21/09/2021
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	23/06/2021	21/09/2021
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	23/06/2021	21/09/2021
					CT321	Long-term surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	23/06/2021	21/09/2021
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	23/06/2021	21/09/2021
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	23/06/2021	21/09/2021
					CT449	Generators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	23/06/2021	21/09/2021
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	23/06/2021	21/09/2021
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	23/06/2021	21/09/2021
					CT981	Single use	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	23/06/2021	21/09/2021
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	23/06/2021	21/09/2021
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	23/06/2021	21/09/2021
					CT983	Surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	23/06/2021	21/09/2021
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	23/06/2021	21/09/2021
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	23/06/2021	21/09/2021
11282	2021	All other reportable incidents	35852	Single-chamber implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem	25/06/2021	25/06/2021
									A070504	Premature Discharge of Battery	25/06/2021	25/06/2021
							F08	Hospitalization or Prolonged Hospitalization	A0705	Battery Problem	25/06/2021	25/06/2021
									A070504	Premature Discharge of Battery	25/06/2021	25/06/2021
							F1905	Device Revision or Replacement	A0705	Battery Problem	25/06/2021	25/06/2021
									A070504	Premature Discharge of Battery	25/06/2021	25/06/2021
11283	2021	All other reportable incidents	47265	Dual-chamber implantable pacemaker, rate-responsive	CT141	Electrodes/Leads and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A0908	Incorrect, Inadequate or Imprecise Result or Readings	21/06/2021	21/06/2021
							F08	Hospitalization or Prolonged Hospitalization	A0908	Incorrect, Inadequate or Imprecise Result or Readings	21/06/2021	21/06/2021
							F1903	Device Explantation	A0908	Incorrect, Inadequate or Imprecise Result or Readings	21/06/2021	21/06/2021
							F1905	Device Revision or Replacement	A0908	Incorrect, Inadequate or Imprecise Result or Readings	21/06/2021	21/06/2021
					CT145	Cardiovascular devices	E2403	No Clinical Signs, Symptoms or Conditions	A0908	Incorrect, Inadequate or Imprecise Result or Readings	21/06/2021	21/06/2021
							F08	Hospitalization or Prolonged Hospitalization	A0908	Incorrect, Inadequate or Imprecise Result or Readings	21/06/2021	21/06/2021
							F1903	Device Explantation	A0908	Incorrect, Inadequate or Imprecise Result or Readings	21/06/2021	21/06/2021
							F1905	Device Revision or Replacement	A0908	Incorrect, Inadequate or Imprecise Result or Readings	21/06/2021	21/06/2021
					CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A0908	Incorrect, Inadequate or Imprecise Result or Readings	21/06/2021	21/06/2021
							F08	Hospitalization or Prolonged Hospitalization	A0908	Incorrect, Inadequate or Imprecise Result or Readings	21/06/2021	21/06/2021
							F1903	Device Explantation	A0908	Incorrect, Inadequate or Imprecise Result or Readings	21/06/2021	21/06/2021
							F1905	Device Revision or Replacement	A0908	Incorrect, Inadequate or Imprecise Result or Readings	21/06/2021	21/06/2021
					CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A0908	Incorrect, Inadequate or Imprecise Result or Readings	21/06/2021	21/06/2021
							F08	Hospitalization or Prolonged Hospitalization	A0908	Incorrect, Inadequate or Imprecise Result or Readings	21/06/2021	21/06/2021
							F1903	Device Explantation	A0908	Incorrect, Inadequate or Imprecise Result or Readings	21/06/2021	21/06/2021
							F1905	Device Revision or Replacement	A0908	Incorrect, Inadequate or Imprecise Result or Readings	21/06/2021	21/06/2021
					CT233	Surgical	E2403	No Clinical Signs, Symptoms or Conditions	A0908	Incorrect, Inadequate or Imprecise Result or Readings	21/06/2021	21/06/2021

							F08	Hospitalization or Prolonged Hospitalization	A0908	Incorrect, Inadequate or Imprecise Result or Readings	21/06/2021	21/06/2021
							F1903	Device Explantation	A0908	Incorrect, Inadequate or Imprecise Result or Readings	21/06/2021	21/06/2021
							F1905	Device Revision or Replacement	A0908	Incorrect, Inadequate or Imprecise Result or Readings	21/06/2021	21/06/2021
					CT2406	Implantable	E2403	No Clinical Signs, Symptoms or Conditions	A0908	Incorrect, Inadequate or Imprecise Result or Readings	21/06/2021	21/06/2021
							F08	Hospitalization or Prolonged Hospitalization	A0908	Incorrect, Inadequate or Imprecise Result or Readings	21/06/2021	21/06/2021
							F1903	Device Explantation	A0908	Incorrect, Inadequate or Imprecise Result or Readings	21/06/2021	21/06/2021
							F1905	Device Revision or Replacement	A0908	Incorrect, Inadequate or Imprecise Result or Readings	21/06/2021	21/06/2021
					CT2407	Active implantable	E2403	No Clinical Signs, Symptoms or Conditions	A0908	Incorrect, Inadequate or Imprecise Result or Readings	21/06/2021	21/06/2021
							F08	Hospitalization or Prolonged Hospitalization	A0908	Incorrect, Inadequate or Imprecise Result or Readings	21/06/2021	21/06/2021
							F1903	Device Explantation	A0908	Incorrect, Inadequate or Imprecise Result or Readings	21/06/2021	21/06/2021
							F1905	Device Revision or Replacement	A0908	Incorrect, Inadequate or Imprecise Result or Readings	21/06/2021	21/06/2021
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A0908	Incorrect, Inadequate or Imprecise Result or Readings	21/06/2021	21/06/2021
							F08	Hospitalization or Prolonged Hospitalization	A0908	Incorrect, Inadequate or Imprecise Result or Readings	21/06/2021	21/06/2021
							F1903	Device Explantation	A0908	Incorrect, Inadequate or Imprecise Result or Readings	21/06/2021	21/06/2021
							F1905	Device Revision or Replacement	A0908	Incorrect, Inadequate or Imprecise Result or Readings	21/06/2021	21/06/2021
					CT2727	Active implant devices/ancillaries	E2403	No Clinical Signs, Symptoms or Conditions	A0908	Incorrect, Inadequate or Imprecise Result or Readings	21/06/2021	21/06/2021
							F08	Hospitalization or Prolonged Hospitalization	A0908	Incorrect, Inadequate or Imprecise Result or Readings	21/06/2021	21/06/2021
							F1903	Device Explantation	A0908	Incorrect, Inadequate or Imprecise Result or Readings	21/06/2021	21/06/2021
							F1905	Device Revision or Replacement	A0908	Incorrect, Inadequate or Imprecise Result or Readings	21/06/2021	21/06/2021
					CT2750	Implantable cardiac electrical stimulators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A0908	Incorrect, Inadequate or Imprecise Result or Readings	21/06/2021	21/06/2021
							F08	Hospitalization or Prolonged Hospitalization	A0908	Incorrect, Inadequate or Imprecise Result or Readings	21/06/2021	21/06/2021
							F1903	Device Explantation	A0908	Incorrect, Inadequate or Imprecise Result or Readings	21/06/2021	21/06/2021
							F1905	Device Revision or Replacement	A0908	Incorrect, Inadequate or Imprecise Result or Readings	21/06/2021	21/06/2021
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A0908	Incorrect, Inadequate or Imprecise Result or Readings	21/06/2021	21/06/2021
							F08	Hospitalization or Prolonged Hospitalization	A0908	Incorrect, Inadequate or Imprecise Result or Readings	21/06/2021	21/06/2021
							F1903	Device Explantation	A0908	Incorrect, Inadequate or Imprecise Result or Readings	21/06/2021	21/06/2021
							F1905	Device Revision or Replacement	A0908	Incorrect, Inadequate or Imprecise Result or Readings	21/06/2021	21/06/2021
					CT2985	Active	E2403	No Clinical Signs, Symptoms or Conditions	A0908	Incorrect, Inadequate or Imprecise Result or Readings	21/06/2021	21/06/2021
							F08	Hospitalization or Prolonged Hospitalization	A0908	Incorrect, Inadequate or Imprecise Result or Readings	21/06/2021	21/06/2021
							F1903	Device Explantation	A0908	Incorrect, Inadequate or Imprecise Result or Readings	21/06/2021	21/06/2021
							F1905	Device Revision or Replacement	A0908	Incorrect, Inadequate or Imprecise Result or Readings	21/06/2021	21/06/2021
					CT3001	Implantable metal	E2403	No Clinical Signs, Symptoms or Conditions	A0908	Incorrect, Inadequate or Imprecise Result or Readings	21/06/2021	21/06/2021
							F08	Hospitalization or Prolonged Hospitalization	A0908	Incorrect, Inadequate or Imprecise Result or Readings	21/06/2021	21/06/2021
							F1903	Device Explantation	A0908	Incorrect, Inadequate or Imprecise Result or Readings	21/06/2021	21/06/2021
							F1905	Device Revision or Replacement	A0908	Incorrect, Inadequate or Imprecise Result or Readings	21/06/2021	21/06/2021
					CT3002	Implantable synthetic polymer	E2403	No Clinical Signs, Symptoms or Conditions	A0908	Incorrect, Inadequate or Imprecise Result or Readings	21/06/2021	21/06/2021
							F08	Hospitalization or Prolonged Hospitalization	A0908	Incorrect, Inadequate or Imprecise Result or Readings	21/06/2021	21/06/2021
							F1903	Device Explantation	A0908	Incorrect, Inadequate or Imprecise Result or Readings	21/06/2021	21/06/2021
							F1905	Device Revision or Replacement	A0908	Incorrect, Inadequate or Imprecise Result or Readings	21/06/2021	21/06/2021
					CT3006	Implantable inorganic material	E2403	No Clinical Signs, Symptoms or Conditions	A0908	Incorrect, Inadequate or Imprecise Result or Readings	21/06/2021	21/06/2021
							F08	Hospitalization or Prolonged Hospitalization	A0908	Incorrect, Inadequate or Imprecise Result or Readings	21/06/2021	21/06/2021
							F1903	Device Explantation	A0908	Incorrect, Inadequate or Imprecise Result or Readings	21/06/2021	21/06/2021
							F1905	Device Revision or Replacement	A0908	Incorrect, Inadequate or Imprecise Result or Readings	21/06/2021	21/06/2021
					CT314	Electrically powered	E2403	No Clinical Signs, Symptoms or Conditions	A0908	Incorrect, Inadequate or Imprecise Result or Readings	21/06/2021	21/06/2021
							F08	Hospitalization or Prolonged Hospitalization	A0908	Incorrect, Inadequate or Imprecise Result or Readings	21/06/2021	21/06/2021
							F1903	Device Explantation	A0908	Incorrect, Inadequate or Imprecise Result or Readings	21/06/2021	21/06/2021
							F1905	Device Revision or Replacement	A0908	Incorrect, Inadequate or Imprecise Result or Readings	21/06/2021	21/06/2021
					CT321	Long-term surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A0908	Incorrect, Inadequate or Imprecise Result or Readings	21/06/2021	21/06/2021
							F08	Hospitalization or Prolonged Hospitalization	A0908	Incorrect, Inadequate or Imprecise Result or Readings	21/06/2021	21/06/2021
							F1903	Device Explantation	A0908	Incorrect, Inadequate or Imprecise Result or Readings	21/06/2021	21/06/2021

						F1905	Device Revision or Replacement	A0908	Incorrect, Inadequate or Imprecise Result or Readings	21/06/2021	21/06/2021
					CT449	Generators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A0908	Incorrect, Inadequate or Imprecise Result or Readings	21/06/2021	21/06/2021
							F08	Hospitalization or Prolonged Hospitalization	A0908	Incorrect, Inadequate or Imprecise Result or Readings	21/06/2021	21/06/2021
							F1903	Device Explantation	A0908	Incorrect, Inadequate or Imprecise Result or Readings	21/06/2021	21/06/2021
							F1905	Device Revision or Replacement	A0908	Incorrect, Inadequate or Imprecise Result or Readings	21/06/2021	21/06/2021
					CT981	Single use	E2403	No Clinical Signs, Symptoms or Conditions	A0908	Incorrect, Inadequate or Imprecise Result or Readings	21/06/2021	21/06/2021
							F08	Hospitalization or Prolonged Hospitalization	A0908	Incorrect, Inadequate or Imprecise Result or Readings	21/06/2021	21/06/2021
							F1903	Device Explantation	A0908	Incorrect, Inadequate or Imprecise Result or Readings	21/06/2021	21/06/2021
							F1905	Device Revision or Replacement	A0908	Incorrect, Inadequate or Imprecise Result or Readings	21/06/2021	21/06/2021
					CT983	Surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A0908	Incorrect, Inadequate or Imprecise Result or Readings	21/06/2021	21/06/2021
							F08	Hospitalization or Prolonged Hospitalization	A0908	Incorrect, Inadequate or Imprecise Result or Readings	21/06/2021	21/06/2021
							F1903	Device Explantation	A0908	Incorrect, Inadequate or Imprecise Result or Readings	21/06/2021	21/06/2021
							F1905	Device Revision or Replacement	A0908	Incorrect, Inadequate or Imprecise Result or Readings	21/06/2021	21/06/2021
11284	2021	All other reportable incidents	37265	Dual-chamber implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E060101	Asystole	A070909	Oversensing		
									A0712	Pacing Problem		
							E0623	Unspecified Heart Problem	A070909	Oversensing		
									A0712	Pacing Problem		
							F05	Delay to Treatment/ Therapy	A070909	Oversensing		
									A0712	Pacing Problem		
							F15	Recognized Device or Procedural Complication	A070909	Oversensing		
									A0712	Pacing Problem		
11285	2021	All other reportable incidents	47267	Single-chamber implantable pacemaker, rate-responsive	CT141	Electrodes/Leads and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	06/09/2021	06/09/2021
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	06/09/2021	06/09/2021
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	06/09/2021	06/09/2021
					CT145	Cardiovascular devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	06/09/2021	06/09/2021
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	06/09/2021	06/09/2021
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	06/09/2021	06/09/2021
					CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	06/09/2021	06/09/2021
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	06/09/2021	06/09/2021
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	06/09/2021	06/09/2021
					CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	06/09/2021	06/09/2021
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	06/09/2021	06/09/2021
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	06/09/2021	06/09/2021
					CT233	Surgical	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	06/09/2021	06/09/2021
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	06/09/2021	06/09/2021
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	06/09/2021	06/09/2021
					CT2406	Implantable	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	06/09/2021	06/09/2021
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	06/09/2021	06/09/2021
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	06/09/2021	06/09/2021
					CT2407	Active implantable	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	06/09/2021	06/09/2021
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	06/09/2021	06/09/2021
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	06/09/2021	06/09/2021
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	06/09/2021	06/09/2021
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	06/09/2021	06/09/2021
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	06/09/2021	06/09/2021
					CT2727	Active implant devices/ancillaries	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	06/09/2021	06/09/2021
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	06/09/2021	06/09/2021
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	06/09/2021	06/09/2021
					CT2750	Implantable cardiac electrical stimulators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	06/09/2021	06/09/2021
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	06/09/2021	06/09/2021
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	06/09/2021	06/09/2021
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	06/09/2021	06/09/2021
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	06/09/2021	06/09/2021
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	06/09/2021	06/09/2021
					CT2985	Active	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	06/09/2021	06/09/2021
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	06/09/2021	06/09/2021
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	06/09/2021	06/09/2021
					CT3001	Implantable metal	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	06/09/2021	06/09/2021
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	06/09/2021	06/09/2021
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	06/09/2021	06/09/2021
					CT3002	Implantable synthetic polymer	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	06/09/2021	06/09/2021
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	06/09/2021	06/09/2021
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	06/09/2021	06/09/2021
					CT3006	Implantable inorganic material	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	06/09/2021	06/09/2021
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	06/09/2021	06/09/2021
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	06/09/2021	06/09/2021
					CT314	Electrically powered	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	06/09/2021	06/09/2021
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	06/09/2021	06/09/2021
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	06/09/2021	06/09/2021
					CT321	Long-term surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	06/09/2021	06/09/2021
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	06/09/2021	06/09/2021
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	06/09/2021	06/09/2021
					CT449	Generators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	06/09/2021	06/09/2021
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	06/09/2021	06/09/2021
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	06/09/2021	06/09/2021
					CT981	Single use	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	06/09/2021	06/09/2021
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	06/09/2021	06/09/2021
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	06/09/2021	06/09/2021

					CT983	Surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	06/09/2021	06/09/2021
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	06/09/2021	06/09/2021
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	06/09/2021	06/09/2021
11286	2021	All other reportable incidents	35223	Endocardial/interventricular septal pacing lead	CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070102	High Capture Threshold		
									A070901	Decreased Sensitivity		
									A072202	Low Impedance		
							F26	No Health Consequences or Impact	A070102	High Capture Threshold		
									A070901	Decreased Sensitivity		
									A072202	Low Impedance		
11287	2021	All other reportable incidents	47270	Cardiac resynchronization therapy implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	01/07/2021	01/07/2021
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery	01/07/2021	01/07/2021
11288	2021	Death	47270	Cardiac resynchronization therapy implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E0119	Loss of Consciousness	A0709	Device Sensing Problem	27/06/2021	27/06/2021
									A070910	Undersensing	27/06/2021	27/06/2021
									A071301	Failure to Deliver Shock/Stimulation	27/06/2021	27/06/2021
							E060109	Tachycardia	A0709	Device Sensing Problem	27/06/2021	27/06/2021
									A070910	Undersensing	27/06/2021	27/06/2021
									A071301	Failure to Deliver Shock/Stimulation	27/06/2021	27/06/2021
							E060110	Ventricular Fibrillation	A0709	Device Sensing Problem	27/06/2021	27/06/2021
									A070910	Undersensing	27/06/2021	27/06/2021
									A071301	Failure to Deliver Shock/Stimulation	27/06/2021	27/06/2021
							E0602	Cardiac Arrest	A0709	Device Sensing Problem	27/06/2021	27/06/2021
									A070910	Undersensing	27/06/2021	27/06/2021
									A071301	Failure to Deliver Shock/Stimulation	27/06/2021	27/06/2021
							F02	Death	A0709	Device Sensing Problem	27/06/2021	27/06/2021
									A070910	Undersensing	27/06/2021	27/06/2021
									A071301	Failure to Deliver Shock/Stimulation	27/06/2021	27/06/2021
							F05	Delay to Treatment/ Therapy	A0709	Device Sensing Problem	27/06/2021	27/06/2021
									A070910	Undersensing	27/06/2021	27/06/2021
									A071301	Failure to Deliver Shock/Stimulation	27/06/2021	27/06/2021
							F1203	Life Threatening Illness or Injury	A0709	Device Sensing Problem	27/06/2021	27/06/2021
									A070910	Undersensing	27/06/2021	27/06/2021
									A071301	Failure to Deliver Shock/Stimulation	27/06/2021	27/06/2021
11289	2021	All other reportable incidents	47263	Cardiac resynchronization therapy implantable pacemaker	CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A0908	Incorrect, Inadequate or Imprecise Result or Readings	17/06/2021	17/06/2021
							F08	Hospitalization or Prolonged Hospitalization	A0908	Incorrect, Inadequate or Imprecise Result or Readings	17/06/2021	17/06/2021
							F1903	Device Explantation	A0908	Incorrect, Inadequate or Imprecise Result or Readings	17/06/2021	17/06/2021
							F1905	Device Revision or Replacement	A0908	Incorrect, Inadequate or Imprecise Result or Readings	17/06/2021	17/06/2021
					CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A0908	Incorrect, Inadequate or Imprecise Result or Readings	17/06/2021	17/06/2021
							F08	Hospitalization or Prolonged Hospitalization	A0908	Incorrect, Inadequate or Imprecise Result or Readings	17/06/2021	17/06/2021
							F1903	Device Explantation	A0908	Incorrect, Inadequate or Imprecise Result or Readings	17/06/2021	17/06/2021
							F1905	Device Revision or Replacement	A0908	Incorrect, Inadequate or Imprecise Result or Readings	17/06/2021	17/06/2021
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A0908	Incorrect, Inadequate or Imprecise Result or Readings	17/06/2021	17/06/2021
							F08	Hospitalization or Prolonged Hospitalization	A0908	Incorrect, Inadequate or Imprecise Result or Readings	17/06/2021	17/06/2021
							F1903	Device Explantation	A0908	Incorrect, Inadequate or Imprecise Result or Readings	17/06/2021	17/06/2021
							F1905	Device Revision or Replacement	A0908	Incorrect, Inadequate or Imprecise Result or Readings	17/06/2021	17/06/2021
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A0908	Incorrect, Inadequate or Imprecise Result or Readings	17/06/2021	17/06/2021
							F08	Hospitalization or Prolonged Hospitalization	A0908	Incorrect, Inadequate or Imprecise Result or Readings	17/06/2021	17/06/2021
							F1903	Device Explantation	A0908	Incorrect, Inadequate or Imprecise Result or Readings	17/06/2021	17/06/2021
							F1905	Device Revision or Replacement	A0908	Incorrect, Inadequate or Imprecise Result or Readings	17/06/2021	17/06/2021
11290	2021	All other reportable incidents	35852	Single-chamber implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
11291	2021	All other reportable incidents	47265	Dual-chamber implantable pacemaker, rate-responsive	CT141	Electrodes/Leads and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A072201	High Impedance		
									A2202	Device Difficult to Program or Calibrate		
							F15	Recognized Device or Procedural Complication	A072201	High Impedance		
									A2202	Device Difficult to Program or Calibrate		
					CT145	Cardiovascular devices	E2403	No Clinical Signs, Symptoms or Conditions	A072201	High Impedance		
									A2202	Device Difficult to Program or Calibrate		
							F15	Recognized Device or Procedural Complication	A072201	High Impedance		
									A2202	Device Difficult to Program or Calibrate		
					CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A072201	High Impedance		
									A2202	Device Difficult to Program or Calibrate		
							F15	Recognized Device or Procedural Complication	A072201	High Impedance		
									A2202	Device Difficult to Program or Calibrate		
					CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A072201	High Impedance		
									A2202	Device Difficult to Program or Calibrate		
							F15	Recognized Device or Procedural Complication	A072201	High Impedance		
									A2202	Device Difficult to Program or Calibrate		
					CT233	Surgical	E2403	No Clinical Signs, Symptoms or Conditions	A072201	High Impedance		
									A2202	Device Difficult to Program or Calibrate		
							F15	Recognized Device or Procedural Complication	A072201	High Impedance		
									A2202	Device Difficult to Program or Calibrate		
					CT2406	Implantable	E2403	No Clinical Signs, Symptoms or Conditions	A072201	High Impedance		
									A2202	Device Difficult to Program or Calibrate		
							F15	Recognized Device or Procedural Complication	A072201	High Impedance		
									A2202	Device Difficult to Program or Calibrate		

					CT2407	Active implantable	E2403	No Clinical Signs, Symptoms or Conditions	A072201	High Impedance		
									A2202	Device Difficult to Program or Calibrate		
							F15	Recognized Device or Procedural Complication	A072201	High Impedance		
									A2202	Device Difficult to Program or Calibrate		
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A072201	High Impedance		
									A2202	Device Difficult to Program or Calibrate		
							F15	Recognized Device or Procedural Complication	A072201	High Impedance		
									A2202	Device Difficult to Program or Calibrate		
					CT2727	Active implant devices/ancillaries	E2403	No Clinical Signs, Symptoms or Conditions	A072201	High Impedance		
									A2202	Device Difficult to Program or Calibrate		
							F15	Recognized Device or Procedural Complication	A072201	High Impedance		
									A2202	Device Difficult to Program or Calibrate		
					CT2750	Implantable cardiac electrical stimulators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A072201	High Impedance		
									A2202	Device Difficult to Program or Calibrate		
							F15	Recognized Device or Procedural Complication	A072201	High Impedance		
									A2202	Device Difficult to Program or Calibrate		
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A072201	High Impedance		
									A2202	Device Difficult to Program or Calibrate		
							F15	Recognized Device or Procedural Complication	A072201	High Impedance		
									A2202	Device Difficult to Program or Calibrate		
					CT2985	Active	E2403	No Clinical Signs, Symptoms or Conditions	A072201	High Impedance		
									A2202	Device Difficult to Program or Calibrate		
							F15	Recognized Device or Procedural Complication	A072201	High Impedance		
									A2202	Device Difficult to Program or Calibrate		
					CT3001	Implantable metal	E2403	No Clinical Signs, Symptoms or Conditions	A072201	High Impedance		
									A2202	Device Difficult to Program or Calibrate		
							F15	Recognized Device or Procedural Complication	A072201	High Impedance		
									A2202	Device Difficult to Program or Calibrate		
					CT3002	Implantable synthetic polymer	E2403	No Clinical Signs, Symptoms or Conditions	A072201	High Impedance		
									A2202	Device Difficult to Program or Calibrate		
							F15	Recognized Device or Procedural Complication	A072201	High Impedance		
									A2202	Device Difficult to Program or Calibrate		
					CT3006	Implantable inorganic material	E2403	No Clinical Signs, Symptoms or Conditions	A072201	High Impedance		
									A2202	Device Difficult to Program or Calibrate		
							F15	Recognized Device or Procedural Complication	A072201	High Impedance		
									A2202	Device Difficult to Program or Calibrate		
					CT314	Electrically powered	E2403	No Clinical Signs, Symptoms or Conditions	A072201	High Impedance		
									A2202	Device Difficult to Program or Calibrate		
							F15	Recognized Device or Procedural Complication	A072201	High Impedance		
									A2202	Device Difficult to Program or Calibrate		
					CT321	Long-term surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A072201	High Impedance		
									A2202	Device Difficult to Program or Calibrate		
							F15	Recognized Device or Procedural Complication	A072201	High Impedance		
									A2202	Device Difficult to Program or Calibrate		
					CT449	Generators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A072201	High Impedance		
									A2202	Device Difficult to Program or Calibrate		
							F15	Recognized Device or Procedural Complication	A072201	High Impedance		
									A2202	Device Difficult to Program or Calibrate		
					CT981	Single use	E2403	No Clinical Signs, Symptoms or Conditions	A072201	High Impedance		
									A2202	Device Difficult to Program or Calibrate		
							F15	Recognized Device or Procedural Complication	A072201	High Impedance		
									A2202	Device Difficult to Program or Calibrate		
					CT983	Surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A072201	High Impedance		
									A2202	Device Difficult to Program or Calibrate		
							F15	Recognized Device or Procedural Complication	A072201	High Impedance		
									A2202	Device Difficult to Program or Calibrate		
11292	2021	All other reportable incidents	47265	Dual-chamber implantable pacemaker, rate-responsive	CT141	Electrodes/Leads and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
					CT145	Cardiovascular devices	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
					CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
					CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
					CT233	Surgical	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
					CT2406	Implantable	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
					CT2407	Active implantable	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		

					CT2727	Active implant devices/ancillaries	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
					CT2750	Implantable cardiac electrical stimulators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
					CT2985	Active	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
					CT3001	Implantable metal	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
					CT3002	Implantable synthetic polymer	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
					CT3006	Implantable inorganic material	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
					CT314	Electrically powered	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
					CT321	Long-term surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
					CT449	Generators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
					CT981	Single use	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
					CT983	Surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
11293	2021	All other reportable incidents	35852	Single-chamber implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A070909	Oversensing		
									A090801	Signal Artifact/Noise		
							F13	Misdiagnosis/ Misclassification	A070909	Oversensing		
									A090801	Signal Artifact/Noise		
11294	2021	All other reportable incidents	47265	Dual-chamber implantable pacemaker, rate-responsive	CT141	Electrodes/Leads and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
					CT145	Cardiovascular devices	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
					CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
					CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
					CT233	Surgical	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
					CT2406	Implantable	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
					CT2407	Active implantable	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
					CT2727	Active implant devices/ancillaries	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		

					CT2750	Implantable cardiac electrical stimulators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
					CT2985	Active	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
					CT3001	Implantable metal	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
					CT3002	Implantable synthetic polymer	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
					CT3006	Implantable inorganic material	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
					CT314	Electrically powered	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
					CT321	Long-term surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
					CT449	Generators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
					CT981	Single use	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
					CT983	Surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
11295	2021	All other reportable incidents	35852	Single-chamber implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A070909	Oversensing		
									A090801	Signal Artifact/Noise		
							F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A070909	Oversensing		
									A090801	Signal Artifact/Noise		
							F13	Misdiagnosis/ Misclassification	A070909	Oversensing		
									A090801	Signal Artifact/Noise		
11296	2021	All other reportable incidents	17882	Physiologic-monitoring defibrillation system	CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070803	Failure to Power Up		
							F27	Problem Identified During Non-Clinical Procedure	A070803	Failure to Power Up		
11297	2021	All other reportable incidents	17882	Physiologic-monitoring defibrillation system	CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A07	Electrical /Electronic Property Problem		
							F27	Problem Identified During Non-Clinical Procedure	A07	Electrical /Electronic Property Problem		
11298	2021	All other reportable incidents	47267	Single-chamber implantable pacemaker, rate-responsive	CT141	Electrodes/Leads and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	01/01/2021	31/12/2021
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery	01/01/2021	31/12/2021
					CT145	Cardiovascular devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	01/01/2021	31/12/2021
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery	01/01/2021	31/12/2021
					CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	01/01/2021	31/12/2021
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery	01/01/2021	31/12/2021
					CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	01/01/2021	31/12/2021
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery	01/01/2021	31/12/2021
					CT233	Surgical	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	01/01/2021	31/12/2021
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery	01/01/2021	31/12/2021
					CT2406	Implantable	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	01/01/2021	31/12/2021
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery	01/01/2021	31/12/2021
					CT2407	Active implantable	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	01/01/2021	31/12/2021
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery	01/01/2021	31/12/2021
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	01/01/2021	31/12/2021
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery	01/01/2021	31/12/2021
					CT2727	Active implant devices/ancillaries	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	01/01/2021	31/12/2021
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery	01/01/2021	31/12/2021
					CT2750	Implantable cardiac electrical stimulators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	01/01/2021	31/12/2021
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery	01/01/2021	31/12/2021
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	01/01/2021	31/12/2021
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery	01/01/2021	31/12/2021
					CT2985	Active	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	01/01/2021	31/12/2021
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery	01/01/2021	31/12/2021
					CT3001	Implantable metal	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	01/01/2021	31/12/2021
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery	01/01/2021	31/12/2021
					CT3002	Implantable synthetic polymer	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	01/01/2021	31/12/2021
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery	01/01/2021	31/12/2021
					CT3006	Implantable inorganic material	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	01/01/2021	31/12/2021
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery	01/01/2021	31/12/2021
					CT314	Electrically powered	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	01/01/2021	31/12/2021
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery	01/01/2021	31/12/2021
					CT321	Long-term surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	01/01/2021	31/12/2021
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery	01/01/2021	31/12/2021

					CT449	Generators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	01/01/2021	31/12/2021
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery	01/01/2021	31/12/2021
					CT981	Single use	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	01/01/2021	31/12/2021
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery	01/01/2021	31/12/2021
					CT983	Surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	01/01/2021	31/12/2021
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery	01/01/2021	31/12/2021
11299	2021	Death	60789	Intracardiac pacemaker	CT130	Catheters and associated devices	E0602	Cardiac Arrest	A24	No Apparent Device Problem	02/07/2021	02/07/2021
							E0604	Cardiac Perforation	A24	No Apparent Device Problem	02/07/2021	02/07/2021
							E0619	Pericardial Effusion	A24	No Apparent Device Problem	02/07/2021	02/07/2021
							E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	02/07/2021	02/07/2021
							F02	Death	A24	No Apparent Device Problem	02/07/2021	02/07/2021
							F1203	Life Threatening Illness or Injury	A24	No Apparent Device Problem	02/07/2021	02/07/2021
							F1901	Additional Surgery	A24	No Apparent Device Problem	02/07/2021	02/07/2021
					CT145	Cardiovascular devices	E0602	Cardiac Arrest	A24	No Apparent Device Problem	02/07/2021	02/07/2021
							E0604	Cardiac Perforation	A24	No Apparent Device Problem	02/07/2021	02/07/2021
							E0619	Pericardial Effusion	A24	No Apparent Device Problem	02/07/2021	02/07/2021
							E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	02/07/2021	02/07/2021
							F02	Death	A24	No Apparent Device Problem	02/07/2021	02/07/2021
							F1203	Life Threatening Illness or Injury	A24	No Apparent Device Problem	02/07/2021	02/07/2021
							F1901	Additional Surgery	A24	No Apparent Device Problem	02/07/2021	02/07/2021
					CT1684	Cardiac pacemakers and associated devices	E0602	Cardiac Arrest	A24	No Apparent Device Problem	02/07/2021	02/07/2021
							E0604	Cardiac Perforation	A24	No Apparent Device Problem	02/07/2021	02/07/2021
							E0619	Pericardial Effusion	A24	No Apparent Device Problem	02/07/2021	02/07/2021
							E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	02/07/2021	02/07/2021
							F02	Death	A24	No Apparent Device Problem	02/07/2021	02/07/2021
							F1203	Life Threatening Illness or Injury	A24	No Apparent Device Problem	02/07/2021	02/07/2021
							F1901	Additional Surgery	A24	No Apparent Device Problem	02/07/2021	02/07/2021
					CT1685	Cardiac pacemakers	E0602	Cardiac Arrest	A24	No Apparent Device Problem	02/07/2021	02/07/2021
							E0604	Cardiac Perforation	A24	No Apparent Device Problem	02/07/2021	02/07/2021
							E0619	Pericardial Effusion	A24	No Apparent Device Problem	02/07/2021	02/07/2021
							E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	02/07/2021	02/07/2021
							F02	Death	A24	No Apparent Device Problem	02/07/2021	02/07/2021
							F1203	Life Threatening Illness or Injury	A24	No Apparent Device Problem	02/07/2021	02/07/2021
							F1901	Additional Surgery	A24	No Apparent Device Problem	02/07/2021	02/07/2021
					CT233	Surgical	E0602	Cardiac Arrest	A24	No Apparent Device Problem	02/07/2021	02/07/2021
							E0604	Cardiac Perforation	A24	No Apparent Device Problem	02/07/2021	02/07/2021
							E0619	Pericardial Effusion	A24	No Apparent Device Problem	02/07/2021	02/07/2021
							E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	02/07/2021	02/07/2021
							F02	Death	A24	No Apparent Device Problem	02/07/2021	02/07/2021
							F1203	Life Threatening Illness or Injury	A24	No Apparent Device Problem	02/07/2021	02/07/2021
							F1901	Additional Surgery	A24	No Apparent Device Problem	02/07/2021	02/07/2021
					CT2406	Implantable	E0602	Cardiac Arrest	A24	No Apparent Device Problem	02/07/2021	02/07/2021
							E0604	Cardiac Perforation	A24	No Apparent Device Problem	02/07/2021	02/07/2021
							E0619	Pericardial Effusion	A24	No Apparent Device Problem	02/07/2021	02/07/2021
							E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	02/07/2021	02/07/2021
							F02	Death	A24	No Apparent Device Problem	02/07/2021	02/07/2021
							F1203	Life Threatening Illness or Injury	A24	No Apparent Device Problem	02/07/2021	02/07/2021
							F1901	Additional Surgery	A24	No Apparent Device Problem	02/07/2021	02/07/2021
					CT2407	Active implantable	E0602	Cardiac Arrest	A24	No Apparent Device Problem	02/07/2021	02/07/2021
							E0604	Cardiac Perforation	A24	No Apparent Device Problem	02/07/2021	02/07/2021
							E0619	Pericardial Effusion	A24	No Apparent Device Problem	02/07/2021	02/07/2021
							E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	02/07/2021	02/07/2021
							F02	Death	A24	No Apparent Device Problem	02/07/2021	02/07/2021
							F1203	Life Threatening Illness or Injury	A24	No Apparent Device Problem	02/07/2021	02/07/2021
							F1901	Additional Surgery	A24	No Apparent Device Problem	02/07/2021	02/07/2021
					CT257	Implantable cardiac pacemakers	E0602	Cardiac Arrest	A24	No Apparent Device Problem	02/07/2021	02/07/2021
							E0604	Cardiac Perforation	A24	No Apparent Device Problem	02/07/2021	02/07/2021
							E0619	Pericardial Effusion	A24	No Apparent Device Problem	02/07/2021	02/07/2021
							E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	02/07/2021	02/07/2021
							F02	Death	A24	No Apparent Device Problem	02/07/2021	02/07/2021
							F1203	Life Threatening Illness or Injury	A24	No Apparent Device Problem	02/07/2021	02/07/2021
							F1901	Additional Surgery	A24	No Apparent Device Problem	02/07/2021	02/07/2021
					CT2727	Active implant devices/ancillaries	E0602	Cardiac Arrest	A24	No Apparent Device Problem	02/07/2021	02/07/2021
							E0604	Cardiac Perforation	A24	No Apparent Device Problem	02/07/2021	02/07/2021
							E0619	Pericardial Effusion	A24	No Apparent Device Problem	02/07/2021	02/07/2021
							E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	02/07/2021	02/07/2021
							F02	Death	A24	No Apparent Device Problem	02/07/2021	02/07/2021
							F1203	Life Threatening Illness or Injury	A24	No Apparent Device Problem	02/07/2021	02/07/2021
							F1901	Additional Surgery	A24	No Apparent Device Problem	02/07/2021	02/07/2021
					CT2750	Implantable cardiac electrical stimulators and associated devices	E0602	Cardiac Arrest	A24	No Apparent Device Problem	02/07/2021	02/07/2021
							E0604	Cardiac Perforation	A24	No Apparent Device Problem	02/07/2021	02/07/2021
							E0619	Pericardial Effusion	A24	No Apparent Device Problem	02/07/2021	02/07/2021
							E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	02/07/2021	02/07/2021
							F02	Death	A24	No Apparent Device Problem	02/07/2021	02/07/2021
							F1203	Life Threatening Illness or Injury	A24	No Apparent Device Problem	02/07/2021	02/07/2021
							F1901	Additional Surgery	A24	No Apparent Device Problem	02/07/2021	02/07/2021
					CT2818	Catheterization/Cannulation/Fluid path devices	E0602	Cardiac Arrest	A24	No Apparent Device Problem	02/07/2021	02/07/2021
							E0604	Cardiac Perforation	A24	No Apparent Device Problem	02/07/2021	02/07/2021
							E0619	Pericardial Effusion	A24	No Apparent Device Problem	02/07/2021	02/07/2021
							E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	02/07/2021	02/07/2021
							F02	Death	A24	No Apparent Device Problem	02/07/2021	02/07/2021
							F1203	Life Threatening Illness or Injury	A24	No Apparent Device Problem	02/07/2021	02/07/2021
							F1901	Additional Surgery	A24	No Apparent Device Problem	02/07/2021	02/07/2021
					CT2866	Implantable cardiac electrical stimulators	E0602	Cardiac Arrest	A24	No Apparent Device Problem	02/07/2021	02/07/2021
							E0604	Cardiac Perforation	A24	No Apparent Device Problem	02/07/2021	02/07/2021
							E0619	Pericardial Effusion	A24	No Apparent Device Problem	02/07/2021	02/07/2021
							E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	02/07/2021	02/07/2021
							F02	Death	A24	No Apparent Device Problem	02/07/2021	02/07/2021

							F1203	Life Threatening Illness or Injury	A24	No Apparent Device Problem	02/07/2021	02/07/2021
							F1901	Additional Surgery	A24	No Apparent Device Problem	02/07/2021	02/07/2021
					CT2985	Active	E0602	Cardiac Arrest	A24	No Apparent Device Problem	02/07/2021	02/07/2021
							E0604	Cardiac Perforation	A24	No Apparent Device Problem	02/07/2021	02/07/2021
							E0619	Pericardial Effusion	A24	No Apparent Device Problem	02/07/2021	02/07/2021
							E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	02/07/2021	02/07/2021
							F02	Death	A24	No Apparent Device Problem	02/07/2021	02/07/2021
							F1203	Life Threatening Illness or Injury	A24	No Apparent Device Problem	02/07/2021	02/07/2021
							F1901	Additional Surgery	A24	No Apparent Device Problem	02/07/2021	02/07/2021
					CT3001	Implantable metal	E0602	Cardiac Arrest	A24	No Apparent Device Problem	02/07/2021	02/07/2021
							E0604	Cardiac Perforation	A24	No Apparent Device Problem	02/07/2021	02/07/2021
							E0619	Pericardial Effusion	A24	No Apparent Device Problem	02/07/2021	02/07/2021
							E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	02/07/2021	02/07/2021
							F02	Death	A24	No Apparent Device Problem	02/07/2021	02/07/2021
							F1203	Life Threatening Illness or Injury	A24	No Apparent Device Problem	02/07/2021	02/07/2021
							F1901	Additional Surgery	A24	No Apparent Device Problem	02/07/2021	02/07/2021
					CT3002	Implantable synthetic polymer	E0602	Cardiac Arrest	A24	No Apparent Device Problem	02/07/2021	02/07/2021
							E0604	Cardiac Perforation	A24	No Apparent Device Problem	02/07/2021	02/07/2021
							E0619	Pericardial Effusion	A24	No Apparent Device Problem	02/07/2021	02/07/2021
							E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	02/07/2021	02/07/2021
							F02	Death	A24	No Apparent Device Problem	02/07/2021	02/07/2021
							F1203	Life Threatening Illness or Injury	A24	No Apparent Device Problem	02/07/2021	02/07/2021
							F1901	Additional Surgery	A24	No Apparent Device Problem	02/07/2021	02/07/2021
					CT3006	Implantable inorganic material	E0602	Cardiac Arrest	A24	No Apparent Device Problem	02/07/2021	02/07/2021
							E0604	Cardiac Perforation	A24	No Apparent Device Problem	02/07/2021	02/07/2021
							E0619	Pericardial Effusion	A24	No Apparent Device Problem	02/07/2021	02/07/2021
							E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	02/07/2021	02/07/2021
							F02	Death	A24	No Apparent Device Problem	02/07/2021	02/07/2021
							F1203	Life Threatening Illness or Injury	A24	No Apparent Device Problem	02/07/2021	02/07/2021
							F1901	Additional Surgery	A24	No Apparent Device Problem	02/07/2021	02/07/2021
					CT314	Electrically powered	E0602	Cardiac Arrest	A24	No Apparent Device Problem	02/07/2021	02/07/2021
							E0604	Cardiac Perforation	A24	No Apparent Device Problem	02/07/2021	02/07/2021
							E0619	Pericardial Effusion	A24	No Apparent Device Problem	02/07/2021	02/07/2021
							E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	02/07/2021	02/07/2021
							F02	Death	A24	No Apparent Device Problem	02/07/2021	02/07/2021
							F1203	Life Threatening Illness or Injury	A24	No Apparent Device Problem	02/07/2021	02/07/2021
							F1901	Additional Surgery	A24	No Apparent Device Problem	02/07/2021	02/07/2021
					CT321	Long-term surgical invasive	E0602	Cardiac Arrest	A24	No Apparent Device Problem	02/07/2021	02/07/2021
							E0604	Cardiac Perforation	A24	No Apparent Device Problem	02/07/2021	02/07/2021
							E0619	Pericardial Effusion	A24	No Apparent Device Problem	02/07/2021	02/07/2021
							E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	02/07/2021	02/07/2021
							F02	Death	A24	No Apparent Device Problem	02/07/2021	02/07/2021
							F1203	Life Threatening Illness or Injury	A24	No Apparent Device Problem	02/07/2021	02/07/2021
							F1901	Additional Surgery	A24	No Apparent Device Problem	02/07/2021	02/07/2021
					CT981	Single use	E0602	Cardiac Arrest	A24	No Apparent Device Problem	02/07/2021	02/07/2021
							E0604	Cardiac Perforation	A24	No Apparent Device Problem	02/07/2021	02/07/2021
							E0619	Pericardial Effusion	A24	No Apparent Device Problem	02/07/2021	02/07/2021
							E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	02/07/2021	02/07/2021
							F02	Death	A24	No Apparent Device Problem	02/07/2021	02/07/2021
							F1203	Life Threatening Illness or Injury	A24	No Apparent Device Problem	02/07/2021	02/07/2021
							F1901	Additional Surgery	A24	No Apparent Device Problem	02/07/2021	02/07/2021
					CT983	Surgical invasive	E0602	Cardiac Arrest	A24	No Apparent Device Problem	02/07/2021	02/07/2021
							E0604	Cardiac Perforation	A24	No Apparent Device Problem	02/07/2021	02/07/2021
							E0619	Pericardial Effusion	A24	No Apparent Device Problem	02/07/2021	02/07/2021
							E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	02/07/2021	02/07/2021
							F02	Death	A24	No Apparent Device Problem	02/07/2021	02/07/2021
							F1203	Life Threatening Illness or Injury	A24	No Apparent Device Problem	02/07/2021	02/07/2021
							F1901	Additional Surgery	A24	No Apparent Device Problem	02/07/2021	02/07/2021
11300	2021	All other reportable incidents	17882	Physiologic-monitoring defibrillation system	CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A1208	Fitting Problem		
							F27	Problem Identified During Non-Clinical Procedure	A1208	Fitting Problem		
11301	2021	All other reportable incidents	60910	Coronary venous pacing lead	CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A05	Mechanical Problem		
									A070102	High Capture Threshold		
									A0722	Impedance Problem		
									A072202	Low Impedance		
									A090801	Signal Artifact/Noise		
							F26	No Health Consequences or Impact	A05	Mechanical Problem		
									A070102	High Capture Threshold		
									A0722	Impedance Problem		
									A072202	Low Impedance		
									A090801	Signal Artifact/Noise		
11302	2021	All other reportable incidents	17882	Physiologic-monitoring defibrillation system	CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070803	Failure to Power Up		
							F26	No Health Consequences or Impact	A070803	Failure to Power Up		
11303	2021	All other reportable incidents	47270	Cardiac resynchronization therapy implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A070909	Oversensing		
									A0712	Pacing Problem		
							F26	No Health Consequences or Impact	A070909	Oversensing		
									A0712	Pacing Problem		
11304	2021	All other reportable incidents	47263	Cardiac resynchronization therapy implantable pacemaker	CT1684	Cardiac pacemakers and associated devices	E2103	Device Overstimulation of Tissue	A070504	Premature Discharge of Battery	23/06/2021	23/06/2021
									A0722	Impedance Problem	23/06/2021	23/06/2021
									A0908	Incorrect, Inadequate or Imprecise Result or Readings	23/06/2021	23/06/2021
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	23/06/2021	23/06/2021
									A0722	Impedance Problem	23/06/2021	23/06/2021
									A0908	Incorrect, Inadequate or Imprecise Result or Readings	23/06/2021	23/06/2021
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	23/06/2021	23/06/2021
									A0722	Impedance Problem	23/06/2021	23/06/2021
									A0908	Incorrect, Inadequate or Imprecise Result or Readings	23/06/2021	23/06/2021

					CT1685	Cardiac pacemakers	E2103	Device Overstimulation of Tissue	A070504	Premature Discharge of Battery	23/06/2021	23/06/2021
									A0722	Impedance Problem	23/06/2021	23/06/2021
									A0908	Incorrect, Inadequate or Imprecise Result or Readings	23/06/2021	23/06/2021
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	23/06/2021	23/06/2021
									A0722	Impedance Problem	23/06/2021	23/06/2021
									A0908	Incorrect, Inadequate or Imprecise Result or Readings	23/06/2021	23/06/2021
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	23/06/2021	23/06/2021
									A0722	Impedance Problem	23/06/2021	23/06/2021
									A0908	Incorrect, Inadequate or Imprecise Result or Readings	23/06/2021	23/06/2021
					CT257	Implantable cardiac pacemakers	E2103	Device Overstimulation of Tissue	A070504	Premature Discharge of Battery	23/06/2021	23/06/2021
									A0722	Impedance Problem	23/06/2021	23/06/2021
									A0908	Incorrect, Inadequate or Imprecise Result or Readings	23/06/2021	23/06/2021
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	23/06/2021	23/06/2021
									A0722	Impedance Problem	23/06/2021	23/06/2021
									A0908	Incorrect, Inadequate or Imprecise Result or Readings	23/06/2021	23/06/2021
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	23/06/2021	23/06/2021
									A0722	Impedance Problem	23/06/2021	23/06/2021
									A0908	Incorrect, Inadequate or Imprecise Result or Readings	23/06/2021	23/06/2021
					CT2866	Implantable cardiac electrical stimulators	E2103	Device Overstimulation of Tissue	A070504	Premature Discharge of Battery	23/06/2021	23/06/2021
									A0722	Impedance Problem	23/06/2021	23/06/2021
									A0908	Incorrect, Inadequate or Imprecise Result or Readings	23/06/2021	23/06/2021
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	23/06/2021	23/06/2021
									A0722	Impedance Problem	23/06/2021	23/06/2021
									A0908	Incorrect, Inadequate or Imprecise Result or Readings	23/06/2021	23/06/2021
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	23/06/2021	23/06/2021
									A0722	Impedance Problem	23/06/2021	23/06/2021
									A0908	Incorrect, Inadequate or Imprecise Result or Readings	23/06/2021	23/06/2021
11305	2020	All other reportable incidents			CT257	Implantable cardiac pacemakers	E233001	Chest Pain	A071209	Extracardiac Stimulation		
							E2403	No Clinical Signs, Symptoms or Conditions	A071209	Extracardiac Stimulation		
							F26	No Health Consequences or Impact	A071209	Extracardiac Stimulation		
11306	2020		47263	Cardiac resynchronization therapy implantable pacemaker	CT257	Implantable cardiac pacemakers						
11307	2022	All other reportable incidents	12913	Pacemaker, implantable	CT145	Cardiovascular devices	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem		
							F26	No Health Consequences or Impact	A0705	Battery Problem		
					CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem		
							F26	No Health Consequences or Impact	A0705	Battery Problem		
					CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem		
							F26	No Health Consequences or Impact	A0705	Battery Problem		
					CT2406	Implantable	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem		
							F26	No Health Consequences or Impact	A0705	Battery Problem		
					CT2407	Active implantable	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem		
							F26	No Health Consequences or Impact	A0705	Battery Problem		
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem		
							F26	No Health Consequences or Impact	A0705	Battery Problem		
					CT2727	Active implant devices/ancillaries	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem		
							F26	No Health Consequences or Impact	A0705	Battery Problem		
					CT2750	Implantable cardiac electrical stimulators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem		
							F26	No Health Consequences or Impact	A0705	Battery Problem		
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem		
							F26	No Health Consequences or Impact	A0705	Battery Problem		
					CT2985	Active	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem		
							F26	No Health Consequences or Impact	A0705	Battery Problem		
					CT314	Electrically powered	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem		
							F26	No Health Consequences or Impact	A0705	Battery Problem		
11308	2022	Serious public health threat	12913	Pacemaker, implantable	CT145	Cardiovascular devices	E2403	No Clinical Signs, Symptoms or Conditions	A071205	Pacemaker Found in Back-Up Mode		
							F26	No Health Consequences or Impact	A071205	Pacemaker Found in Back-Up Mode		
					CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A071205	Pacemaker Found in Back-Up Mode		
							F26	No Health Consequences or Impact	A071205	Pacemaker Found in Back-Up Mode		
					CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A071205	Pacemaker Found in Back-Up Mode		
							F26	No Health Consequences or Impact	A071205	Pacemaker Found in Back-Up Mode		
					CT2406	Implantable	E2403	No Clinical Signs, Symptoms or Conditions	A071205	Pacemaker Found in Back-Up Mode		
							F26	No Health Consequences or Impact	A071205	Pacemaker Found in Back-Up Mode		
					CT2407	Active implantable	E2403	No Clinical Signs, Symptoms or Conditions	A071205	Pacemaker Found in Back-Up Mode		
							F26	No Health Consequences or Impact	A071205	Pacemaker Found in Back-Up Mode		
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A071205	Pacemaker Found in Back-Up Mode		
							F26	No Health Consequences or Impact	A071205	Pacemaker Found in Back-Up Mode		
					CT2727	Active implant devices/ancillaries	E2403	No Clinical Signs, Symptoms or Conditions	A071205	Pacemaker Found in Back-Up Mode		
							F26	No Health Consequences or Impact	A071205	Pacemaker Found in Back-Up Mode		
					CT2750	Implantable cardiac electrical stimulators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A071205	Pacemaker Found in Back-Up Mode		
							F26	No Health Consequences or Impact	A071205	Pacemaker Found in Back-Up Mode		
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A071205	Pacemaker Found in Back-Up Mode		
							F26	No Health Consequences or Impact	A071205	Pacemaker Found in Back-Up Mode		
					CT2985	Active	E2403	No Clinical Signs, Symptoms or Conditions	A071205	Pacemaker Found in Back-Up Mode		
							F26	No Health Consequences or Impact	A071205	Pacemaker Found in Back-Up Mode		
					CT314	Electrically powered	E2403	No Clinical Signs, Symptoms or Conditions	A071205	Pacemaker Found in Back-Up Mode		
							F26	No Health Consequences or Impact	A071205	Pacemaker Found in Back-Up Mode		
11309	2022	All other reportable incidents			CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem		
							F26	No Health Consequences or Impact	A0705	Battery Problem		
11310	2021				CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions				
11311	2022	All other reportable incidents			CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		

					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
11312	2022		47265	Dual-chamber implantable pacemaker, rate-responsive	CT141	Electrodes/Leads and associated devices	E011903	Syncope/Fainting	A070101	Failure to Capture	20/07/2022	20/07/2022
							E0601	Arrhythmia	A070101	Failure to Capture	20/07/2022	20/07/2022
							F1901	Additional Surgery	A070101	Failure to Capture	20/07/2022	20/07/2022
					CT145	Cardiovascular devices	E011903	Syncope/Fainting	A070101	Failure to Capture	20/07/2022	20/07/2022
							E0601	Arrhythmia	A070101	Failure to Capture	20/07/2022	20/07/2022
							F1901	Additional Surgery	A070101	Failure to Capture	20/07/2022	20/07/2022
					CT1684	Cardiac pacemakers and associated devices	E011903	Syncope/Fainting	A070101	Failure to Capture	20/07/2022	20/07/2022
							E0601	Arrhythmia	A070101	Failure to Capture	20/07/2022	20/07/2022
							F1901	Additional Surgery	A070101	Failure to Capture	20/07/2022	20/07/2022
					CT1685	Cardiac pacemakers	E011903	Syncope/Fainting	A070101	Failure to Capture	20/07/2022	20/07/2022
							E0601	Arrhythmia	A070101	Failure to Capture	20/07/2022	20/07/2022
							F1901	Additional Surgery	A070101	Failure to Capture	20/07/2022	20/07/2022
					CT233	Surgical	E011903	Syncope/Fainting	A070101	Failure to Capture	20/07/2022	20/07/2022
							E0601	Arrhythmia	A070101	Failure to Capture	20/07/2022	20/07/2022
							F1901	Additional Surgery	A070101	Failure to Capture	20/07/2022	20/07/2022
					CT2406	Implantable	E011903	Syncope/Fainting	A070101	Failure to Capture	20/07/2022	20/07/2022
							E0601	Arrhythmia	A070101	Failure to Capture	20/07/2022	20/07/2022
							F1901	Additional Surgery	A070101	Failure to Capture	20/07/2022	20/07/2022
					CT2407	Active implantable	E011903	Syncope/Fainting	A070101	Failure to Capture	20/07/2022	20/07/2022
							E0601	Arrhythmia	A070101	Failure to Capture	20/07/2022	20/07/2022
							F1901	Additional Surgery	A070101	Failure to Capture	20/07/2022	20/07/2022
					CT257	Implantable cardiac pacemakers	E011903	Syncope/Fainting	A070101	Failure to Capture	20/07/2022	20/07/2022
							E0601	Arrhythmia	A070101	Failure to Capture	20/07/2022	20/07/2022
							F1901	Additional Surgery	A070101	Failure to Capture	20/07/2022	20/07/2022
					CT2727	Active implant devices/ancillaries	E011903	Syncope/Fainting	A070101	Failure to Capture	20/07/2022	20/07/2022
							E0601	Arrhythmia	A070101	Failure to Capture	20/07/2022	20/07/2022
							F1901	Additional Surgery	A070101	Failure to Capture	20/07/2022	20/07/2022
					CT2750	Implantable cardiac electrical stimulators and associated devices	E011903	Syncope/Fainting	A070101	Failure to Capture	20/07/2022	20/07/2022
							E0601	Arrhythmia	A070101	Failure to Capture	20/07/2022	20/07/2022
							F1901	Additional Surgery	A070101	Failure to Capture	20/07/2022	20/07/2022
					CT2866	Implantable cardiac electrical stimulators	E011903	Syncope/Fainting	A070101	Failure to Capture	20/07/2022	20/07/2022
							E0601	Arrhythmia	A070101	Failure to Capture	20/07/2022	20/07/2022
							F1901	Additional Surgery	A070101	Failure to Capture	20/07/2022	20/07/2022
					CT2985	Active	E011903	Syncope/Fainting	A070101	Failure to Capture	20/07/2022	20/07/2022
							E0601	Arrhythmia	A070101	Failure to Capture	20/07/2022	20/07/2022
							F1901	Additional Surgery	A070101	Failure to Capture	20/07/2022	20/07/2022
					CT3001	Implantable metal	E011903	Syncope/Fainting	A070101	Failure to Capture	20/07/2022	20/07/2022
							E0601	Arrhythmia	A070101	Failure to Capture	20/07/2022	20/07/2022
							F1901	Additional Surgery	A070101	Failure to Capture	20/07/2022	20/07/2022
					CT3002	Implantable synthetic polymer	E011903	Syncope/Fainting	A070101	Failure to Capture	20/07/2022	20/07/2022
							E0601	Arrhythmia	A070101	Failure to Capture	20/07/2022	20/07/2022
							F1901	Additional Surgery	A070101	Failure to Capture	20/07/2022	20/07/2022
					CT3006	Implantable inorganic material	E011903	Syncope/Fainting	A070101	Failure to Capture	20/07/2022	20/07/2022
							E0601	Arrhythmia	A070101	Failure to Capture	20/07/2022	20/07/2022
							F1901	Additional Surgery	A070101	Failure to Capture	20/07/2022	20/07/2022
					CT314	Electrically powered	E011903	Syncope/Fainting	A070101	Failure to Capture	20/07/2022	20/07/2022
							E0601	Arrhythmia	A070101	Failure to Capture	20/07/2022	20/07/2022
							F1901	Additional Surgery	A070101	Failure to Capture	20/07/2022	20/07/2022
					CT321	Long-term surgical invasive	E011903	Syncope/Fainting	A070101	Failure to Capture	20/07/2022	20/07/2022
							E0601	Arrhythmia	A070101	Failure to Capture	20/07/2022	20/07/2022
							F1901	Additional Surgery	A070101	Failure to Capture	20/07/2022	20/07/2022
					CT449	Generators and associated devices	E011903	Syncope/Fainting	A070101	Failure to Capture	20/07/2022	20/07/2022
							E0601	Arrhythmia	A070101	Failure to Capture	20/07/2022	20/07/2022
							F1901	Additional Surgery	A070101	Failure to Capture	20/07/2022	20/07/2022
					CT981	Single use	E011903	Syncope/Fainting	A070101	Failure to Capture	20/07/2022	20/07/2022
							E0601	Arrhythmia	A070101	Failure to Capture	20/07/2022	20/07/2022
							F1901	Additional Surgery	A070101	Failure to Capture	20/07/2022	20/07/2022
					CT983	Surgical invasive	E011903	Syncope/Fainting	A070101	Failure to Capture	20/07/2022	20/07/2022
							E0601	Arrhythmia	A070101	Failure to Capture	20/07/2022	20/07/2022
							F1901	Additional Surgery	A070101	Failure to Capture	20/07/2022	20/07/2022
11313	2022	All other reportable incidents			CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem		
							F26	No Health Consequences or Impact	A0705	Battery Problem		
11314	2022	All other reportable incidents			CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem		
							F26	No Health Consequences or Impact	A0705	Battery Problem		
11315	2022	All other reportable incidents			CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem		
							F22	Unexpected Diagnostic Intervention	A0705	Battery Problem		
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem		
							F22	Unexpected Diagnostic Intervention	A0705	Battery Problem		
11316	2022	All other reportable incidents			CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A160503	Inappropriate or Unexpected Reset		
							F26	No Health Consequences or Impact	A160503	Inappropriate or Unexpected Reset		
11317	2022	All other reportable incidents	35223	Endocardial/interventricular septal pacing lead	CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070909	Oversensing		
							F1901	Additional Surgery	A070909	Oversensing		
11318	2022	All other reportable incidents	35223	Endocardial/interventricular septal pacing lead	CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070909	Oversensing		
									A0722	Impedance Problem		
									A072201	High Impedance		
							F08	Hospitalization or Prolonged Hospitalization	A070909	Oversensing		
									A0722	Impedance Problem		
									A072201	High Impedance		
							F12	Serious Injury/ Illness/ Impairment	A070909	Oversensing		
									A0722	Impedance Problem		
									A072201	High Impedance		
							F23	Unexpected Medical Intervention	A070909	Oversensing		
									A0722	Impedance Problem		
									A072201	High Impedance		
11319	2021	All other reportable incidents			CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	26/01/2022	26/01/2022
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	26/01/2022	26/01/2022

							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	26/01/2022	26/01/2022
11320	2021	All other reportable incidents			CT257	Implantable cardiac pacemakers	E0601	Arrhythmia	A070910	Undersensing		
									A071201	Failure to Convert Rhythm		
							E2403	No Clinical Signs, Symptoms or Conditions	A070910	Undersensing		
									A071201	Failure to Convert Rhythm		
							F08	Hospitalization or Prolonged Hospitalization	A070910	Undersensing		
									A071201	Failure to Convert Rhythm		
							F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A070910	Undersensing		
									A071201	Failure to Convert Rhythm		
11321	2021	All other reportable incidents	35223	Endocardial/interventricular septal pacing lead	CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070909	Oversensing		
									A071204	No Pacing		
							F23	Unexpected Medical Intervention	A070909	Oversensing		
									A071204	No Pacing		
11322	2021	All other reportable incidents			CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	13/12/2021	13/12/2021
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery	13/12/2021	13/12/2021
11323	2021	All other reportable incidents	35223	Endocardial/interventricular septal pacing lead	CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A0722	Impedance Problem		
							F26	No Health Consequences or Impact	A0722	Impedance Problem		
11324	2022		35223	Endocardial/interventricular septal pacing lead	CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070909	Oversensing		
							F26	No Health Consequences or Impact	A070909	Oversensing		
11325	2021	All other reportable incidents	47270	Cardiac resynchronization therapy implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A0908	Incorrect, Inadequate or Imprecise Result or Readings	27/11/2021	27/11/2021
							F08	Hospitalization or Prolonged Hospitalization	A0908	Incorrect, Inadequate or Imprecise Result or Readings	27/11/2021	27/11/2021
							F1905	Device Revision or Replacement	A0908	Incorrect, Inadequate or Imprecise Result or Readings	27/11/2021	27/11/2021
11326	2021	All other reportable incidents	35852	Single-chamber implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem	24/01/2022	24/01/2022
									A070504	Premature Discharge of Battery	24/01/2022	24/01/2022
							F08	Hospitalization or Prolonged Hospitalization	A0705	Battery Problem	24/01/2022	24/01/2022
									A070504	Premature Discharge of Battery	24/01/2022	24/01/2022
							F1905	Device Revision or Replacement	A0705	Battery Problem	24/01/2022	24/01/2022
									A070504	Premature Discharge of Battery	24/01/2022	24/01/2022
11327	2021	All other reportable incidents			CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070803	Failure to Power Up		
							F27	Problem Identified During Non-Clinical Procedure	A070803	Failure to Power Up		
11328	2021	All other reportable incidents	47263	Cardiac resynchronization therapy implantable pacemaker	CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A150206	Difficult to Insert	29/11/2021	29/11/2021
							F1908	Prolonged Surgery	A150206	Difficult to Insert	29/11/2021	29/11/2021
					CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A150206	Difficult to Insert	29/11/2021	29/11/2021
							F1908	Prolonged Surgery	A150206	Difficult to Insert	29/11/2021	29/11/2021
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A150206	Difficult to Insert	29/11/2021	29/11/2021
							F1908	Prolonged Surgery	A150206	Difficult to Insert	29/11/2021	29/11/2021
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A150206	Difficult to Insert	29/11/2021	29/11/2021
							F1908	Prolonged Surgery	A150206	Difficult to Insert	29/11/2021	29/11/2021
11329	2022	All other reportable incidents	47267	Single-chamber implantable pacemaker, rate-responsive	CT141	Electrodes/Leads and associated devices	F15	Recognized Device or Procedural Complication	A0705	Battery Problem		
					CT145	Cardiovascular devices	F15	Recognized Device or Procedural Complication	A0705	Battery Problem		
					CT1684	Cardiac pacemakers and associated devices	F15	Recognized Device or Procedural Complication	A0705	Battery Problem		
					CT1685	Cardiac pacemakers	F15	Recognized Device or Procedural Complication	A0705	Battery Problem		
					CT233	Surgical	F15	Recognized Device or Procedural Complication	A0705	Battery Problem		
					CT2406	Implantable	F15	Recognized Device or Procedural Complication	A0705	Battery Problem		
					CT2407	Active implantable	F15	Recognized Device or Procedural Complication	A0705	Battery Problem		
					CT257	Implantable cardiac pacemakers	F15	Recognized Device or Procedural Complication	A0705	Battery Problem		
					CT2727	Active implant devices/ancillaries	F15	Recognized Device or Procedural Complication	A0705	Battery Problem		
					CT2750	Implantable cardiac electrical stimulators and associated devices	F15	Recognized Device or Procedural Complication	A0705	Battery Problem		
					CT2866	Implantable cardiac electrical stimulators	F15	Recognized Device or Procedural Complication	A0705	Battery Problem		
					CT2985	Active	F15	Recognized Device or Procedural Complication	A0705	Battery Problem		
					CT3001	Implantable metal	F15	Recognized Device or Procedural Complication	A0705	Battery Problem		
					CT3002	Implantable synthetic polymer	F15	Recognized Device or Procedural Complication	A0705	Battery Problem		
					CT3006	Implantable inorganic material	F15	Recognized Device or Procedural Complication	A0705	Battery Problem		
					CT314	Electrically powered	F15	Recognized Device or Procedural Complication	A0705	Battery Problem		
					CT321	Long-term surgical invasive	F15	Recognized Device or Procedural Complication	A0705	Battery Problem		
					CT449	Generators and associated devices	F15	Recognized Device or Procedural Complication	A0705	Battery Problem		
					CT981	Single use	F15	Recognized Device or Procedural Complication	A0705	Battery Problem		
					CT983	Surgical invasive	F15	Recognized Device or Procedural Complication	A0705	Battery Problem		
11330	2022	All other reportable incidents			CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions				
							F26	No Health Consequences or Impact				
11331	2022	All other reportable incidents			CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem		
							F23	Unexpected Medical Intervention	A0705	Battery Problem		
11332	2021	All other reportable incidents			CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A07	Electrical /Electronic Property Problem		
							F27	Problem Identified During Non-Clinical Procedure	A07	Electrical /Electronic Property Problem		
11333	2022	All other reportable incidents			CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A0908	Incorrect, Inadequate or Imprecise Result or Readings	26/07/2022	26/07/2022
							F08	Hospitalization or Prolonged Hospitalization	A0908	Incorrect, Inadequate or Imprecise Result or Readings	26/07/2022	26/07/2022
							F1905	Device Revision or Replacement	A0908	Incorrect, Inadequate or Imprecise Result or Readings	26/07/2022	26/07/2022
11334	2022	Death			CT257	Implantable cardiac pacemakers	E0602	Cardiac Arrest	A09	Output Problem		
							F02	Death	A09	Output Problem		
11335	2022	All other reportable incidents			CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A07	Electrical /Electronic Property Problem		
							F26	No Health Consequences or Impact	A07	Electrical /Electronic Property Problem		
11336	2022	All other reportable incidents			CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A07	Electrical /Electronic Property Problem		
							F26	No Health Consequences or Impact	A07	Electrical /Electronic Property Problem		
11337	2022	All other reportable incidents	47267	Single-chamber implantable pacemaker, rate-responsive	CT141	Electrodes/Leads and associated devices	F15	Recognized Device or Procedural Complication	A05	Mechanical Problem		
									A0705	Battery Problem		
					CT145	Cardiovascular devices	F15	Recognized Device or Procedural Complication	A05	Mechanical Problem		
									A0705	Battery Problem		
					CT1684	Cardiac pacemakers and associated devices	F15	Recognized Device or Procedural Complication	A05	Mechanical Problem		
									A0705	Battery Problem		
					CT1685	Cardiac pacemakers	F15	Recognized Device or Procedural Complication	A05	Mechanical Problem		
									A0705	Battery Problem		
					CT233	Surgical	F15	Recognized Device or Procedural Complication	A05	Mechanical Problem		

									A0705	Battery Problem		
					CT2406	Implantable	F15	Recognized Device or Procedural Complication	A05	Mechanical Problem		
									A0705	Battery Problem		
					CT2407	Active implantable	F15	Recognized Device or Procedural Complication	A05	Mechanical Problem		
									A0705	Battery Problem		
					CT257	Implantable cardiac pacemakers	F15	Recognized Device or Procedural Complication	A05	Mechanical Problem		
									A0705	Battery Problem		
					CT2727	Active implant devices/ancillaries	F15	Recognized Device or Procedural Complication	A05	Mechanical Problem		
									A0705	Battery Problem		
					CT2750	Implantable cardiac electrical stimulators and associated devices	F15	Recognized Device or Procedural Complication	A05	Mechanical Problem		
									A0705	Battery Problem		
					CT2866	Implantable cardiac electrical stimulators	F15	Recognized Device or Procedural Complication	A05	Mechanical Problem		
									A0705	Battery Problem		
					CT2985	Active	F15	Recognized Device or Procedural Complication	A05	Mechanical Problem		
									A0705	Battery Problem		
					CT3001	Implantable metal	F15	Recognized Device or Procedural Complication	A05	Mechanical Problem		
									A0705	Battery Problem		
					CT3002	Implantable synthetic polymer	F15	Recognized Device or Procedural Complication	A05	Mechanical Problem		
									A0705	Battery Problem		
					CT3006	Implantable inorganic material	F15	Recognized Device or Procedural Complication	A05	Mechanical Problem		
									A0705	Battery Problem		
					CT314	Electrically powered	F15	Recognized Device or Procedural Complication	A05	Mechanical Problem		
									A0705	Battery Problem		
					CT321	Long-term surgical invasive	F15	Recognized Device or Procedural Complication	A05	Mechanical Problem		
									A0705	Battery Problem		
					CT449	Generators and associated devices	F15	Recognized Device or Procedural Complication	A05	Mechanical Problem		
									A0705	Battery Problem		
					CT981	Single use	F15	Recognized Device or Procedural Complication	A05	Mechanical Problem		
									A0705	Battery Problem		
					CT983	Surgical invasive	F15	Recognized Device or Procedural Complication	A05	Mechanical Problem		
									A0705	Battery Problem		
11338	2022	All other reportable incidents	35223	Endocardial/interventricular septal pacing lead	CT1684	Cardiac pacemakers and associated devices	F01	Change in Therapeutic Response	A040102	Loss of or Failure to Bond		
11339	2022	All other reportable incidents	35223	Endocardial/interventricular septal pacing lead	CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A160601	Premature Elective Replacement Indicator		
							F01	Change in Therapeutic Response	A160601	Premature Elective Replacement Indicator		
11340	2022		35223	Endocardial/interventricular septal pacing lead	CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070909	Oversensing		
							F23	Unexpected Medical Intervention	A070909	Oversensing		
11341	2022	All other reportable incidents	17882	Physiologic-monitoring defibrillation system	CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A0404	Crack		
							F05	Delay to Treatment/ Therapy	A0404	Crack		
							F26	No Health Consequences or Impact	A0404	Crack		
11342	2022	All other reportable incidents	47270	Cardiac resynchronization therapy implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem		
							F2301	Additional Device Required	A0705	Battery Problem		
11343	2022	All other reportable incidents	47265	Dual-chamber implantable pacemaker, rate-responsive	CT141	Electrodes/Leads and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT145	Cardiovascular devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT233	Surgical	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT2406	Implantable	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT2407	Active implantable	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT2727	Active implant devices/ancillaries	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT2750	Implantable cardiac electrical stimulators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT2985	Active	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT3001	Implantable metal	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT3002	Implantable synthetic polymer	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT3006	Implantable inorganic material	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT314	Electrically powered	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT321	Long-term surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT449	Generators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT981	Single use	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT983	Surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
11344	2022	All other reportable incidents			CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A090203	Erratic or Intermittent Display		
							F26	No Health Consequences or Impact	A090203	Erratic or Intermittent Display		
11345	2022	All other reportable incidents	17882	Physiologic-monitoring defibrillation system	CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions				
									A0706	Charging Problem		
							F26	No Health Consequences or Impact				
									A0706	Charging Problem		

11346	2019	Unanticipated serious deterioration in state of health			CT257	Implantable cardiac pacemakers	E1008	Diarrhea	A010101	Biocompatibility	06/04/2019	06/04/2019
									A0906	Gas Output Problem	06/04/2019	06/04/2019
							E131001	Kidney Infection	A010101	Biocompatibility	06/04/2019	06/04/2019
									A0906	Gas Output Problem	06/04/2019	06/04/2019
							F0101	Therapeutic Response Decreased	A010101	Biocompatibility	06/04/2019	06/04/2019
									A0906	Gas Output Problem	06/04/2019	06/04/2019
							F13	Misdiagnosis/ Misclassification	A010101	Biocompatibility	06/04/2019	06/04/2019
									A0906	Gas Output Problem	06/04/2019	06/04/2019
11347	2022	All other reportable incidents	35223	Endocardial/interventricular septal pacing lead	CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A090801	Signal Artifact/Noise		
							F26	No Health Consequences or Impact	A090801	Signal Artifact/Noise		
11348	2022	Death	17882	Physiologic-monitoring defibrillation system	CT1684	Cardiac pacemakers and associated devices	E0602	Cardiac Arrest	A11	Computer Software Problem		
							F02	Death	A11	Computer Software Problem		
11349	2022	All other reportable incidents			CT257	Implantable cardiac pacemakers	E2103	Device Overstimulation of Tissue	A071205	Pacemaker Found in Back-Up Mode		
							F1905	Device Revision or Replacement	A071205	Pacemaker Found in Back-Up Mode		
							F26	No Health Consequences or Impact	A071205	Pacemaker Found in Back-Up Mode		
11350	2022	All other reportable incidents			CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A071101	Difficult to Interrogate		
									A1104	Computer Operating System Problem		
							F26	No Health Consequences or Impact	A071101	Difficult to Interrogate		
									A1104	Computer Operating System Problem		
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A071101	Difficult to Interrogate		
									A1104	Computer Operating System Problem		
							F26	No Health Consequences or Impact	A071101	Difficult to Interrogate		
									A1104	Computer Operating System Problem		
11351	2022	All other reportable incidents	12913	Pacemaker, implantable	CT145	Cardiovascular devices	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem		
							F1905	Device Revision or Replacement	A0705	Battery Problem		
					CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem		
							F1905	Device Revision or Replacement	A0705	Battery Problem		
					CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem		
							F1905	Device Revision or Replacement	A0705	Battery Problem		
					CT2406	Implantable	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem		
							F1905	Device Revision or Replacement	A0705	Battery Problem		
					CT2407	Active implantable	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem		
							F1905	Device Revision or Replacement	A0705	Battery Problem		
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem		
							F1905	Device Revision or Replacement	A0705	Battery Problem		
					CT2727	Active implant devices/ancillaries	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem		
							F1905	Device Revision or Replacement	A0705	Battery Problem		
					CT2750	Implantable cardiac electrical stimulators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem		
							F1905	Device Revision or Replacement	A0705	Battery Problem		
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem		
							F1905	Device Revision or Replacement	A0705	Battery Problem		
					CT2985	Active	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem		
							F1905	Device Revision or Replacement	A0705	Battery Problem		
					CT314	Electrically powered	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem		
							F1905	Device Revision or Replacement	A0705	Battery Problem		
11352	2022	All other reportable incidents			CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A0203	Defective Device		
							F1905	Device Revision or Replacement	A0203	Defective Device		
11353	2022	All other reportable incidents	37265	Dual-chamber implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	F19	Surgical Intervention	A0705	Battery Problem		
11354	2022	Unanticipated serious deterioration in state of health	60789	Intracardiac pacemaker	CT130	Catheters and associated devices			A070910	Undersensing		
					CT145	Cardiovascular devices			A070910	Undersensing		
					CT1684	Cardiac pacemakers and associated devices			A070910	Undersensing		
					CT1685	Cardiac pacemakers			A070910	Undersensing		
					CT233	Surgical			A070910	Undersensing		
					CT2406	Implantable			A070910	Undersensing		
					CT2407	Active implantable			A070910	Undersensing		
					CT257	Implantable cardiac pacemakers			A070910	Undersensing		
					CT2727	Active implant devices/ancillaries			A070910	Undersensing		
					CT2750	Implantable cardiac electrical stimulators and associated devices			A070910	Undersensing		
					CT2818	Catheterization/Cannulation/Fluid path devices			A070910	Undersensing		
					CT2866	Implantable cardiac electrical stimulators			A070910	Undersensing		
					CT2985	Active			A070910	Undersensing		
					CT3001	Implantable metal			A070910	Undersensing		
					CT3002	Implantable synthetic polymer			A070910	Undersensing		
					CT3006	Implantable inorganic material			A070910	Undersensing		
					CT314	Electrically powered			A070910	Undersensing		
					CT321	Long-term surgical invasive			A070910	Undersensing		
					CT981	Single use			A070910	Undersensing		
					CT983	Surgical invasive			A070910	Undersensing		
11355	2022		47265	Dual-chamber implantable pacemaker, rate-responsive	CT141	Electrodes/Leads and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	26/08/2022	26/08/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	26/08/2022	26/08/2022
					CT145	Cardiovascular devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	26/08/2022	26/08/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	26/08/2022	26/08/2022
					CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	26/08/2022	26/08/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	26/08/2022	26/08/2022
					CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	26/08/2022	26/08/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	26/08/2022	26/08/2022
					CT233	Surgical	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	26/08/2022	26/08/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	26/08/2022	26/08/2022
					CT2406	Implantable	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	26/08/2022	26/08/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	26/08/2022	26/08/2022
					CT2407	Active implantable	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	26/08/2022	26/08/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	26/08/2022	26/08/2022
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	26/08/2022	26/08/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	26/08/2022	26/08/2022
					CT2727	Active implant devices/ancillaries	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	26/08/2022	26/08/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	26/08/2022	26/08/2022
					CT2750	Implantable cardiac electrical stimulators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	26/08/2022	26/08/2022

							F1901	Additional Surgery	A24	No Apparent Device Problem	26/08/2022	26/08/2022
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	26/08/2022	26/08/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	26/08/2022	26/08/2022
					CT2985	Active	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	26/08/2022	26/08/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	26/08/2022	26/08/2022
					CT3001	Implantable metal	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	26/08/2022	26/08/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	26/08/2022	26/08/2022
					CT3002	Implantable synthetic polymer	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	26/08/2022	26/08/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	26/08/2022	26/08/2022
					CT3006	Implantable inorganic material	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	26/08/2022	26/08/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	26/08/2022	26/08/2022
					CT314	Electrically powered	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	26/08/2022	26/08/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	26/08/2022	26/08/2022
					CT321	Long-term surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	26/08/2022	26/08/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	26/08/2022	26/08/2022
					CT449	Generators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	26/08/2022	26/08/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	26/08/2022	26/08/2022
					CT981	Single use	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	26/08/2022	26/08/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	26/08/2022	26/08/2022
					CT983	Surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	26/08/2022	26/08/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	26/08/2022	26/08/2022
11356	2022	All other reportable incidents	35223	Endocardial/interventricular septal pacing lead	CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A072201	High Impedance		
							F26	No Health Consequences or Impact	A072201	High Impedance		
11357	2022	All other reportable incidents			CT257	Implantable cardiac pacemakers	E1614	Joint Dislocation	A0203	Defective Device		
							F11	Non-Serious Injury/ Illness/ Impairment	A0203	Defective Device		
11358	2022		47265	Dual-chamber implantable pacemaker, rate-responsive	CT141	Electrodes/Leads and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	22/08/2022	22/08/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	22/08/2022	22/08/2022
					CT145	Cardiovascular devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	22/08/2022	22/08/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	22/08/2022	22/08/2022
					CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	22/08/2022	22/08/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	22/08/2022	22/08/2022
					CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	22/08/2022	22/08/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	22/08/2022	22/08/2022
					CT233	Surgical	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	22/08/2022	22/08/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	22/08/2022	22/08/2022
					CT2406	Implantable	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	22/08/2022	22/08/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	22/08/2022	22/08/2022
					CT2407	Active implantable	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	22/08/2022	22/08/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	22/08/2022	22/08/2022
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	22/08/2022	22/08/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	22/08/2022	22/08/2022
					CT2727	Active implant devices/ancillaries	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	22/08/2022	22/08/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	22/08/2022	22/08/2022
					CT2750	Implantable cardiac electrical stimulators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	22/08/2022	22/08/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	22/08/2022	22/08/2022
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	22/08/2022	22/08/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	22/08/2022	22/08/2022
					CT2985	Active	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	22/08/2022	22/08/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	22/08/2022	22/08/2022
					CT3001	Implantable metal	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	22/08/2022	22/08/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	22/08/2022	22/08/2022
					CT3002	Implantable synthetic polymer	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	22/08/2022	22/08/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	22/08/2022	22/08/2022
					CT3006	Implantable inorganic material	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	22/08/2022	22/08/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	22/08/2022	22/08/2022
					CT314	Electrically powered	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	22/08/2022	22/08/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	22/08/2022	22/08/2022
					CT321	Long-term surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	22/08/2022	22/08/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	22/08/2022	22/08/2022
					CT449	Generators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	22/08/2022	22/08/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	22/08/2022	22/08/2022
					CT981	Single use	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	22/08/2022	22/08/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	22/08/2022	22/08/2022
					CT983	Surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	22/08/2022	22/08/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	22/08/2022	22/08/2022
11359	2022	All other reportable incidents	47270	Cardiac resynchronization therapy implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A071101	Difficult to Interrogate		
							F26	No Health Consequences or Impact	A071101	Difficult to Interrogate		
11360	2022	All other reportable incidents			CT257	Implantable cardiac pacemakers	F1907	More Complex Surgery	A0202	Defective Component		
11361	2022	All other reportable incidents			CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem		
							F26	No Health Consequences or Impact	A0705	Battery Problem		
11362	2022	All other reportable incidents			CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem		
							F1905	Device Revision or Replacement	A0705	Battery Problem		
11363	2022	All other reportable incidents	47263	Cardiac resynchronization therapy implantable pacemaker	CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem		
							F23	Unexpected Medical Intervention	A0705	Battery Problem		
					CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem		
							F23	Unexpected Medical Intervention	A0705	Battery Problem		
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem		
							F23	Unexpected Medical Intervention	A0705	Battery Problem		
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem		
							F23	Unexpected Medical Intervention	A0705	Battery Problem		
11364	2022	Unanticipated serious deterioration in state of health	35223	Endocardial/interventricular septal pacing lead	CT1684	Cardiac pacemakers and associated devices	F1904	Device Repositioning	A150201	Positioning Failure		
11365	2022	All other reportable incidents	47270	Cardiac resynchronization therapy implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E0601	Arrhythmia	A070101	Failure to Capture		
									A071204	No Pacing		
							F12	Serious Injury/ Illness/ Impairment	A070101	Failure to Capture		
									A071204	No Pacing		
11366	2022	All other reportable incidents	35852	Single-chamber implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A07	Electrical /Electronic Property Problem		
							F26	No Health Consequences or Impact	A07	Electrical /Electronic Property Problem		

11367	2022	Unanticipated serious deterioration in state of health			CT257	Implantable cardiac pacemakers	E0717	Dyspnea	A071102	Failure to Interrogate		
							F08	Hospitalization or Prolonged Hospitalization	A071102	Failure to Interrogate		
11368	2022	All other reportable incidents	47265	Dual-chamber implantable pacemaker, rate-responsive	CT141	Electrodes/Leads and associated devices	E233001	Chest Pain	A090811	Unable to Obtain Readings	23/08/2022	
							F08	Hospitalization or Prolonged Hospitalization	A090811	Unable to Obtain Readings	23/08/2022	
					CT145	Cardiovascular devices	E233001	Chest Pain	A090811	Unable to Obtain Readings	23/08/2022	
							F08	Hospitalization or Prolonged Hospitalization	A090811	Unable to Obtain Readings	23/08/2022	
					CT1684	Cardiac pacemakers and associated devices	E233001	Chest Pain	A090811	Unable to Obtain Readings	23/08/2022	
							F08	Hospitalization or Prolonged Hospitalization	A090811	Unable to Obtain Readings	23/08/2022	
					CT1685	Cardiac pacemakers	E233001	Chest Pain	A090811	Unable to Obtain Readings	23/08/2022	
							F08	Hospitalization or Prolonged Hospitalization	A090811	Unable to Obtain Readings	23/08/2022	
					CT233	Surgical	E233001	Chest Pain	A090811	Unable to Obtain Readings	23/08/2022	
							F08	Hospitalization or Prolonged Hospitalization	A090811	Unable to Obtain Readings	23/08/2022	
					CT2406	Implantable	E233001	Chest Pain	A090811	Unable to Obtain Readings	23/08/2022	
							F08	Hospitalization or Prolonged Hospitalization	A090811	Unable to Obtain Readings	23/08/2022	
					CT2407	Active implantable	E233001	Chest Pain	A090811	Unable to Obtain Readings	23/08/2022	
							F08	Hospitalization or Prolonged Hospitalization	A090811	Unable to Obtain Readings	23/08/2022	
					CT257	Implantable cardiac pacemakers	E233001	Chest Pain	A090811	Unable to Obtain Readings	23/08/2022	
							F08	Hospitalization or Prolonged Hospitalization	A090811	Unable to Obtain Readings	23/08/2022	
					CT2727	Active implant devices/ancillaries	E233001	Chest Pain	A090811	Unable to Obtain Readings	23/08/2022	
							F08	Hospitalization or Prolonged Hospitalization	A090811	Unable to Obtain Readings	23/08/2022	
					CT2750	Implantable cardiac electrical stimulators and associated devices	E233001	Chest Pain	A090811	Unable to Obtain Readings	23/08/2022	
							F08	Hospitalization or Prolonged Hospitalization	A090811	Unable to Obtain Readings	23/08/2022	
					CT2866	Implantable cardiac electrical stimulators	E233001	Chest Pain	A090811	Unable to Obtain Readings	23/08/2022	
							F08	Hospitalization or Prolonged Hospitalization	A090811	Unable to Obtain Readings	23/08/2022	
					CT2985	Active	E233001	Chest Pain	A090811	Unable to Obtain Readings	23/08/2022	
							F08	Hospitalization or Prolonged Hospitalization	A090811	Unable to Obtain Readings	23/08/2022	
					CT3001	Implantable metal	E233001	Chest Pain	A090811	Unable to Obtain Readings	23/08/2022	
							F08	Hospitalization or Prolonged Hospitalization	A090811	Unable to Obtain Readings	23/08/2022	
					CT3002	Implantable synthetic polymer	E233001	Chest Pain	A090811	Unable to Obtain Readings	23/08/2022	
							F08	Hospitalization or Prolonged Hospitalization	A090811	Unable to Obtain Readings	23/08/2022	
					CT3006	Implantable inorganic material	E233001	Chest Pain	A090811	Unable to Obtain Readings	23/08/2022	
							F08	Hospitalization or Prolonged Hospitalization	A090811	Unable to Obtain Readings	23/08/2022	
					CT314	Electrically powered	E233001	Chest Pain	A090811	Unable to Obtain Readings	23/08/2022	
							F08	Hospitalization or Prolonged Hospitalization	A090811	Unable to Obtain Readings	23/08/2022	
					CT321	Long-term surgical invasive	E233001	Chest Pain	A090811	Unable to Obtain Readings	23/08/2022	
							F08	Hospitalization or Prolonged Hospitalization	A090811	Unable to Obtain Readings	23/08/2022	
					CT449	Generators and associated devices	E233001	Chest Pain	A090811	Unable to Obtain Readings	23/08/2022	
							F08	Hospitalization or Prolonged Hospitalization	A090811	Unable to Obtain Readings	23/08/2022	
					CT981	Single use	E233001	Chest Pain	A090811	Unable to Obtain Readings	23/08/2022	
							F08	Hospitalization or Prolonged Hospitalization	A090811	Unable to Obtain Readings	23/08/2022	
					CT983	Surgical invasive	E233001	Chest Pain	A090811	Unable to Obtain Readings	23/08/2022	
							F08	Hospitalization or Prolonged Hospitalization	A090811	Unable to Obtain Readings	23/08/2022	
11369	2022	All other reportable incidents	35223	Endocardial/interventricular septal pacing lead	CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A0401	Break		
									A040101	Fracture		
									A070909	Oversensing		
									A0722	Impedance Problem		
									A090801	Signal Artifact/Noise		
							F26	No Health Consequences or Impact	A0401	Break		
									A040101	Fracture		
									A070909	Oversensing		
									A0722	Impedance Problem		
									A090801	Signal Artifact/Noise		
11370	2022	All other reportable incidents	35223	Endocardial/interventricular septal pacing lead	CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions				
							F26	No Health Consequences or Impact				
11371	2022	All other reportable incidents	35852	Single-chamber implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F28	Appropriate Health Impact Term/Code not Available	A070504	Premature Discharge of Battery		
11372	2022	All other reportable incidents	35852	Single-chamber implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	F1905	Device Revision or Replacement	A0705	Battery Problem		
11373	2022	All other reportable incidents			CT257	Implantable cardiac pacemakers	F1905	Device Revision or Replacement	A0705	Battery Problem		
11374	2022	All other reportable incidents	35223	Endocardial/interventricular septal pacing lead	CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A072202	Low Impedance		
							F1905	Device Revision or Replacement	A072202	Low Impedance		
11375	2022	All other reportable incidents	35852	Single-chamber implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A010402	Migration		
							F1905	Device Revision or Replacement	A010402	Migration		
11376	2022	All other reportable incidents	35223	Endocardial/interventricular septal pacing lead	CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A072202	Low Impedance		
							F1905	Device Revision or Replacement	A072202	Low Impedance		
11377	2022	All other reportable incidents	35852	Single-chamber implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem		
									A071205	Pacemaker Found in Back-Up Mode		
							F26	No Health Consequences or Impact	A0705	Battery Problem		
									A071205	Pacemaker Found in Back-Up Mode		
11378	2022	All other reportable incidents			CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A0203	Defective Device		
							F27	Problem Identified During Non-Clinical Procedure	A0203	Defective Device		
11379	2022	All other reportable incidents			CT257	Implantable cardiac pacemakers	F1905	Device Revision or Replacement	A0705	Battery Problem		
11380	2022	All other reportable incidents	47263	Cardiac resynchronization therapy implantable pacemaker	CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A0711	Interrogation Problem	01/07/2022	01/07/2022
									A071102	Failure to Interrogate	01/07/2022	01/07/2022
							F1901	Additional Surgery	A0711	Interrogation Problem	01/07/2022	01/07/2022
									A071102	Failure to Interrogate	01/07/2022	01/07/2022
					CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A0711	Interrogation Problem	01/07/2022	01/07/2022
									A071102	Failure to Interrogate	01/07/2022	01/07/2022
							F1901	Additional Surgery	A0711	Interrogation Problem	01/07/2022	01/07/2022
									A071102	Failure to Interrogate	01/07/2022	01/07/2022
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A0711	Interrogation Problem	01/07/2022	01/07/2022
									A071102	Failure to Interrogate	01/07/2022	01/07/2022
							F1901	Additional Surgery	A0711	Interrogation Problem	01/07/2022	01/07/2022
									A071102	Failure to Interrogate	01/07/2022	01/07/2022
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A0711	Interrogation Problem	01/07/2022	01/07/2022
									A071102	Failure to Interrogate	01/07/2022	01/07/2022
							F1901	Additional Surgery	A0711	Interrogation Problem	01/07/2022	01/07/2022
									A071102	Failure to Interrogate	01/07/2022	01/07/2022
11381	2022	All other reportable incidents	47263	Cardiac resynchronization therapy implantable	CT1684	Cardiac pacemakers and associated devices	F1905	Device Revision or Replacement	A0705	Battery Problem		

				pacemaker	CT1685	Cardiac pacemakers	F1905	Device Revision or Replacement	A0705	Battery Problem		
					CT257	Implantable cardiac pacemakers	F1905	Device Revision or Replacement	A0705	Battery Problem		
					CT2866	Implantable cardiac electrical stimulators	F1905	Device Revision or Replacement	A0705	Battery Problem		
11382	2022	All other reportable incidents	60789	Intracardiac pacemaker	CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A0101	Patient-Device Incompatibility		
							F26	No Health Consequences or Impact	A0101	Patient-Device Incompatibility		
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A0101	Patient-Device Incompatibility		
							F26	No Health Consequences or Impact	A0101	Patient-Device Incompatibility		
					CT2818	Catheterization/Cannulation/Fluid path devices	E2403	No Clinical Signs, Symptoms or Conditions	A0101	Patient-Device Incompatibility		
							F26	No Health Consequences or Impact	A0101	Patient-Device Incompatibility		
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A0101	Patient-Device Incompatibility		
							F26	No Health Consequences or Impact	A0101	Patient-Device Incompatibility		
11383	2022		35223	Endocardial/interventricular septal pacing lead	CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A0701	Capturing Problem	27/04/2022	27/04/2022
									A0709	Device Sensing Problem	27/04/2022	27/04/2022
							F1901	Additional Surgery	A0701	Capturing Problem	27/04/2022	27/04/2022
									A0709	Device Sensing Problem	27/04/2022	27/04/2022
11384	2022	Unanticipated serious deterioration in state of health	60789	Intracardiac pacemaker	CT130	Catheters and associated devices			A27	Appropriate Device Problem Term/Code not Available		
					CT145	Cardiovascular devices			A27	Appropriate Device Problem Term/Code not Available		
					CT1684	Cardiac pacemakers and associated devices			A27	Appropriate Device Problem Term/Code not Available		
					CT1685	Cardiac pacemakers			A27	Appropriate Device Problem Term/Code not Available		
					CT233	Surgical			A27	Appropriate Device Problem Term/Code not Available		
					CT2406	Implantable			A27	Appropriate Device Problem Term/Code not Available		
					CT2407	Active implantable			A27	Appropriate Device Problem Term/Code not Available		
					CT257	Implantable cardiac pacemakers			A27	Appropriate Device Problem Term/Code not Available		
					CT2727	Active implant devices/ancillaries			A27	Appropriate Device Problem Term/Code not Available		
					CT2750	Implantable cardiac electrical stimulators and associated devices			A27	Appropriate Device Problem Term/Code not Available		
					CT2818	Catheterization/Cannulation/Fluid path devices			A27	Appropriate Device Problem Term/Code not Available		
					CT2866	Implantable cardiac electrical stimulators			A27	Appropriate Device Problem Term/Code not Available		
					CT2985	Active			A27	Appropriate Device Problem Term/Code not Available		
					CT3001	Implantable metal			A27	Appropriate Device Problem Term/Code not Available		
					CT3002	Implantable synthetic polymer			A27	Appropriate Device Problem Term/Code not Available		
					CT3006	Implantable inorganic material			A27	Appropriate Device Problem Term/Code not Available		
					CT314	Electrically powered			A27	Appropriate Device Problem Term/Code not Available		
					CT321	Long-term surgical invasive			A27	Appropriate Device Problem Term/Code not Available		
					CT981	Single use			A27	Appropriate Device Problem Term/Code not Available		
					CT983	Surgical invasive			A27	Appropriate Device Problem Term/Code not Available		
11385	2022	All other reportable incidents			CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
11386	2022	All other reportable incidents	37265	Dual-chamber implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem		
							F26	No Health Consequences or Impact	A0705	Battery Problem		
11387	2022	All other reportable incidents	47270	Cardiac resynchronization therapy implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E0138	Undesired Nerve Stimulation	A071205	Pacemaker Found in Back-Up Mode		
							F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A071205	Pacemaker Found in Back-Up Mode		
							F11	Non-Serious Injury/ Illness/ Impairment	A071205	Pacemaker Found in Back-Up Mode		
11388	2022	All other reportable incidents	47270	Cardiac resynchronization therapy implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A0713	Defibrillation/Stimulation Problem		
							F26	No Health Consequences or Impact	A0713	Defibrillation/Stimulation Problem		
11389	2022		47205	Cardiac pulse generator programmer	CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A090805	Incorrect Measurement		
							F23	Unexpected Medical Intervention	A090805	Incorrect Measurement		
11390	2022	All other reportable incidents	35852	Single-chamber implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A160601	Premature Elective Replacement Indicator		
							F26	No Health Consequences or Impact	A160601	Premature Elective Replacement Indicator		
11391	2022	Unanticipated serious deterioration in state of health	35852	Single-chamber implantable defibrillator	CT2866	Implantable cardiac electrical stimulators						
11392	2022	All other reportable incidents	47263	Cardiac resynchronization therapy implantable pacemaker	CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A0203	Defective Device		
							F26	No Health Consequences or Impact	A0203	Defective Device		
					CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A0203	Defective Device		
							F26	No Health Consequences or Impact	A0203	Defective Device		
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A0203	Defective Device		
							F26	No Health Consequences or Impact	A0203	Defective Device		
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A0203	Defective Device		
							F26	No Health Consequences or Impact	A0203	Defective Device		
11393	2022	All other reportable incidents	47263	Cardiac resynchronization therapy implantable pacemaker	CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
11394	2022	All other reportable incidents	35852	Single-chamber implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem	18/08/2022	18/08/2022
									A070504	Premature Discharge of Battery	18/08/2022	18/08/2022
							F08	Hospitalization or Prolonged Hospitalization	A0705	Battery Problem	18/08/2022	18/08/2022
									A070504	Premature Discharge of Battery	18/08/2022	18/08/2022
							F1905	Device Revision or Replacement	A0705	Battery Problem	18/08/2022	18/08/2022
									A070504	Premature Discharge of Battery	18/08/2022	18/08/2022
11395	2022	All other reportable incidents			CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	12/07/2022	12/07/2022

							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	12/07/2022	12/07/2022
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	12/07/2022	12/07/2022
11396	2022	All other reportable incidents	35852	Single-chamber implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A1415	Air/Gas in Device		
							F26	No Health Consequences or Impact	A1415	Air/Gas in Device		
11397	2022	All other reportable incidents			CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	31/03/2022	31/03/2022
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	31/03/2022	31/03/2022
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	31/03/2022	31/03/2022
11398	2022	All other reportable incidents	35852	Single-chamber implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A0202	Defective Component		
							F19	Surgical Intervention	A0202	Defective Component		
11399	2022	All other reportable incidents			CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem	01/09/2022	01/09/2022
									A070504	Premature Discharge of Battery	01/09/2022	01/09/2022
							F08	Hospitalization or Prolonged Hospitalization	A0705	Battery Problem	01/09/2022	01/09/2022
									A070504	Premature Discharge of Battery	01/09/2022	01/09/2022
							F1905	Device Revision or Replacement	A0705	Battery Problem	01/09/2022	01/09/2022
									A070504	Premature Discharge of Battery	01/09/2022	01/09/2022
11400	2022	All other reportable incidents	35852	Single-chamber implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A0101	Patient-Device Incompatibility		
							F26	No Health Consequences or Impact	A0101	Patient-Device Incompatibility		
11401	2022	All other reportable incidents	35852	Single-chamber implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem	10/10/2022	10/10/2022
									A070504	Premature Discharge of Battery	10/10/2022	10/10/2022
							F08	Hospitalization or Prolonged Hospitalization	A0705	Battery Problem	10/10/2022	10/10/2022
									A070504	Premature Discharge of Battery	10/10/2022	10/10/2022
							F1903	Device Explantation	A0705	Battery Problem	10/10/2022	10/10/2022
									A070504	Premature Discharge of Battery	10/10/2022	10/10/2022
11402	2022	All other reportable incidents			CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A0203	Defective Device		
							F1908	Prolonged Surgery	A0203	Defective Device		
11403	2022	All other reportable incidents	35852	Single-chamber implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem		
							F19	Surgical Intervention	A0705	Battery Problem		
11404	2022	All other reportable incidents	35223	Endocardial/interventricular septal pacing lead	CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A040101	Fracture		
									A071204	No Pacing		
							F1905	Device Revision or Replacement	A040101	Fracture		
									A071204	No Pacing		
11405	2022	All other reportable incidents			CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem		
							F26	No Health Consequences or Impact	A0705	Battery Problem		
11406	2022	All other reportable incidents	35852	Single-chamber implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E2104	Electric Shock	A071302	Inappropriate/Inadequate Shock/Stimulation		
							E2403	No Clinical Signs, Symptoms or Conditions	A071302	Inappropriate/Inadequate Shock/Stimulation		
							F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A071302	Inappropriate/Inadequate Shock/Stimulation		
11407	2022	All other reportable incidents			CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem		
							F26	No Health Consequences or Impact	A0705	Battery Problem		
11408	2022	All other reportable incidents	63189	Implantable pacemaker battery	CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem		
							F1905	Device Revision or Replacement	A0705	Battery Problem		
11409	2022	All other reportable incidents			CT257	Implantable cardiac pacemakers	E0601	Arrhythmia	A0705	Battery Problem		
							E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem		
							F01	Change in Therapeutic Response	A0705	Battery Problem		
11410	2022	All other reportable incidents	35852	Single-chamber implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem		
							F26	No Health Consequences or Impact	A0705	Battery Problem		
11411	2022		47270	Cardiac resynchronization therapy implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	05/10/2022	05/10/2022
							F1901	Additional Surgery	A070504	Premature Discharge of Battery	05/10/2022	05/10/2022
11412	2022		47265	Dual-chamber implantable pacemaker, rate-responsive	CT141	Electrodes/Leads and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	05/10/2022	05/10/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	05/10/2022	05/10/2022
					CT145	Cardiovascular devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	05/10/2022	05/10/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	05/10/2022	05/10/2022
					CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	05/10/2022	05/10/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	05/10/2022	05/10/2022
					CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	05/10/2022	05/10/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	05/10/2022	05/10/2022
					CT233	Surgical	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	05/10/2022	05/10/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	05/10/2022	05/10/2022
					CT2406	Implantable	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	05/10/2022	05/10/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	05/10/2022	05/10/2022
					CT2407	Active implantable	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	05/10/2022	05/10/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	05/10/2022	05/10/2022
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	05/10/2022	05/10/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	05/10/2022	05/10/2022
					CT2727	Active implant devices/ancillaries	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	05/10/2022	05/10/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	05/10/2022	05/10/2022
					CT2750	Implantable cardiac electrical stimulators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	05/10/2022	05/10/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	05/10/2022	05/10/2022
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	05/10/2022	05/10/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	05/10/2022	05/10/2022
					CT2985	Active	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	05/10/2022	05/10/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	05/10/2022	05/10/2022
					CT3001	Implantable metal	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	05/10/2022	05/10/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	05/10/2022	05/10/2022
					CT3002	Implantable synthetic polymer	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	05/10/2022	05/10/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	05/10/2022	05/10/2022
					CT3006	Implantable inorganic material	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	05/10/2022	05/10/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	05/10/2022	05/10/2022
					CT314	Electrically powered	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	05/10/2022	05/10/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	05/10/2022	05/10/2022
					CT321	Long-term surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	05/10/2022	05/10/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	05/10/2022	05/10/2022
					CT449	Generators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	05/10/2022	05/10/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	05/10/2022	05/10/2022
					CT981	Single use	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	05/10/2022	05/10/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	05/10/2022	05/10/2022
					CT983	Surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	05/10/2022	05/10/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	05/10/2022	05/10/2022
11413	2022				CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	28/09/2022	28/09/2022

							F1901	Additional Surgery	A24	No Apparent Device Problem	28/09/2022	28/09/2022
11414	2022	Unanticipated serious deterioration in state of health	35852	Single-chamber implantable defibrillator	CT2866	Implantable cardiac electrical stimulators			A27	Appropriate Device Problem Term/Code not Available		
11415	2022	Unanticipated serious deterioration in state of health	60789	Intracardiac pacemaker	CT130	Catheters and associated devices			A0101	Patient-Device Incompatibility		
					CT145	Cardiovascular devices			A0101	Patient-Device Incompatibility		
					CT1684	Cardiac pacemakers and associated devices			A0101	Patient-Device Incompatibility		
					CT1685	Cardiac pacemakers			A0101	Patient-Device Incompatibility		
					CT233	Surgical			A0101	Patient-Device Incompatibility		
					CT2406	Implantable			A0101	Patient-Device Incompatibility		
					CT2407	Active implantable			A0101	Patient-Device Incompatibility		
					CT257	Implantable cardiac pacemakers			A0101	Patient-Device Incompatibility		
					CT2727	Active implant devices/ancillaries			A0101	Patient-Device Incompatibility		
					CT2750	Implantable cardiac electrical stimulators and associated devices			A0101	Patient-Device Incompatibility		
					CT2818	Catheterization/Cannulation/Fluid path devices			A0101	Patient-Device Incompatibility		
					CT2866	Implantable cardiac electrical stimulators			A0101	Patient-Device Incompatibility		
					CT2985	Active			A0101	Patient-Device Incompatibility		
					CT3001	Implantable metal			A0101	Patient-Device Incompatibility		
					CT3002	Implantable synthetic polymer			A0101	Patient-Device Incompatibility		
					CT3006	Implantable inorganic material			A0101	Patient-Device Incompatibility		
					CT314	Electrically powered			A0101	Patient-Device Incompatibility		
					CT321	Long-term surgical invasive			A0101	Patient-Device Incompatibility		
					CT981	Single use			A0101	Patient-Device Incompatibility		
					CT983	Surgical invasive			A0101	Patient-Device Incompatibility		
11416	2022	All other reportable incidents	37265	Dual-chamber implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E2104	Electric Shock	A0712	Pacing Problem		
									A071302	Inappropriate/Inadequate Shock/Stimulation		
							F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A0712	Pacing Problem		
									A071302	Inappropriate/Inadequate Shock/Stimulation		
11417	2022	All other reportable incidents	47267	Single-chamber implantable pacemaker, rate-responsive	CT141	Electrodes/Leads and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem	29/11/2022	29/11/2022
									A070504	Premature Discharge of Battery	29/11/2022	29/11/2022
							F08	Hospitalization or Prolonged Hospitalization	A0705	Battery Problem	29/11/2022	29/11/2022
									A070504	Premature Discharge of Battery	29/11/2022	29/11/2022
							F1905	Device Revision or Replacement	A0705	Battery Problem	29/11/2022	29/11/2022
									A070504	Premature Discharge of Battery	29/11/2022	29/11/2022
					CT145	Cardiovascular devices	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem	29/11/2022	29/11/2022
									A070504	Premature Discharge of Battery	29/11/2022	29/11/2022
							F08	Hospitalization or Prolonged Hospitalization	A0705	Battery Problem	29/11/2022	29/11/2022
									A070504	Premature Discharge of Battery	29/11/2022	29/11/2022
							F1905	Device Revision or Replacement	A0705	Battery Problem	29/11/2022	29/11/2022
									A070504	Premature Discharge of Battery	29/11/2022	29/11/2022
					CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem	29/11/2022	29/11/2022
									A070504	Premature Discharge of Battery	29/11/2022	29/11/2022
							F08	Hospitalization or Prolonged Hospitalization	A0705	Battery Problem	29/11/2022	29/11/2022
									A070504	Premature Discharge of Battery	29/11/2022	29/11/2022
							F1905	Device Revision or Replacement	A0705	Battery Problem	29/11/2022	29/11/2022
									A070504	Premature Discharge of Battery	29/11/2022	29/11/2022
					CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem	29/11/2022	29/11/2022
									A070504	Premature Discharge of Battery	29/11/2022	29/11/2022
							F08	Hospitalization or Prolonged Hospitalization	A0705	Battery Problem	29/11/2022	29/11/2022
									A070504	Premature Discharge of Battery	29/11/2022	29/11/2022
							F1905	Device Revision or Replacement	A0705	Battery Problem	29/11/2022	29/11/2022
									A070504	Premature Discharge of Battery	29/11/2022	29/11/2022
					CT233	Surgical	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem	29/11/2022	29/11/2022
									A070504	Premature Discharge of Battery	29/11/2022	29/11/2022
							F08	Hospitalization or Prolonged Hospitalization	A0705	Battery Problem	29/11/2022	29/11/2022
									A070504	Premature Discharge of Battery	29/11/2022	29/11/2022
							F1905	Device Revision or Replacement	A0705	Battery Problem	29/11/2022	29/11/2022
									A070504	Premature Discharge of Battery	29/11/2022	29/11/2022
					CT2406	Implantable	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem	29/11/2022	29/11/2022
									A070504	Premature Discharge of Battery	29/11/2022	29/11/2022
							F08	Hospitalization or Prolonged Hospitalization	A0705	Battery Problem	29/11/2022	29/11/2022
									A070504	Premature Discharge of Battery	29/11/2022	29/11/2022
							F1905	Device Revision or Replacement	A0705	Battery Problem	29/11/2022	29/11/2022
									A070504	Premature Discharge of Battery	29/11/2022	29/11/2022
					CT2407	Active implantable	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem	29/11/2022	29/11/2022
									A070504	Premature Discharge of Battery	29/11/2022	29/11/2022
							F08	Hospitalization or Prolonged Hospitalization	A0705	Battery Problem	29/11/2022	29/11/2022
									A070504	Premature Discharge of Battery	29/11/2022	29/11/2022
							F1905	Device Revision or Replacement	A0705	Battery Problem	29/11/2022	29/11/2022
									A070504	Premature Discharge of Battery	29/11/2022	29/11/2022
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem	29/11/2022	29/11/2022
									A070504	Premature Discharge of Battery	29/11/2022	29/11/2022
							F08	Hospitalization or Prolonged Hospitalization	A0705	Battery Problem	29/11/2022	29/11/2022
									A070504	Premature Discharge of Battery	29/11/2022	29/11/2022
							F1905	Device Revision or Replacement	A0705	Battery Problem	29/11/2022	29/11/2022
									A070504	Premature Discharge of Battery	29/11/2022	29/11/2022
					CT2727	Active implant devices/ancillaries	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem	29/11/2022	29/11/2022
									A070504	Premature Discharge of Battery	29/11/2022	29/11/2022
							F08	Hospitalization or Prolonged Hospitalization	A0705	Battery Problem	29/11/2022	29/11/2022
									A070504	Premature Discharge of Battery	29/11/2022	29/11/2022
							F1905	Device Revision or Replacement	A0705	Battery Problem	29/11/2022	29/11/2022
									A070504	Premature Discharge of Battery	29/11/2022	29/11/2022
					CT2750	Implantable cardiac electrical stimulators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem	29/11/2022	29/11/2022
									A070504	Premature Discharge of Battery	29/11/2022	29/11/2022
							F08	Hospitalization or Prolonged Hospitalization	A0705	Battery Problem	29/11/2022	29/11/2022
									A070504	Premature Discharge of Battery	29/11/2022	29/11/2022
							F1905	Device Revision or Replacement	A0705	Battery Problem	29/11/2022	29/11/2022
									A070504	Premature Discharge of Battery	29/11/2022	29/11/2022

					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem	29/11/2022	29/11/2022
									A070504	Premature Discharge of Battery	29/11/2022	29/11/2022
							F08	Hospitalization or Prolonged Hospitalization	A0705	Battery Problem	29/11/2022	29/11/2022
									A070504	Premature Discharge of Battery	29/11/2022	29/11/2022
							F1905	Device Revision or Replacement	A0705	Battery Problem	29/11/2022	29/11/2022
									A070504	Premature Discharge of Battery	29/11/2022	29/11/2022
					CT2985	Active	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem	29/11/2022	29/11/2022
									A070504	Premature Discharge of Battery	29/11/2022	29/11/2022
							F08	Hospitalization or Prolonged Hospitalization	A0705	Battery Problem	29/11/2022	29/11/2022
									A070504	Premature Discharge of Battery	29/11/2022	29/11/2022
							F1905	Device Revision or Replacement	A0705	Battery Problem	29/11/2022	29/11/2022
									A070504	Premature Discharge of Battery	29/11/2022	29/11/2022
					CT3001	Implantable metal	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem	29/11/2022	29/11/2022
									A070504	Premature Discharge of Battery	29/11/2022	29/11/2022
							F08	Hospitalization or Prolonged Hospitalization	A0705	Battery Problem	29/11/2022	29/11/2022
									A070504	Premature Discharge of Battery	29/11/2022	29/11/2022
							F1905	Device Revision or Replacement	A0705	Battery Problem	29/11/2022	29/11/2022
									A070504	Premature Discharge of Battery	29/11/2022	29/11/2022
					CT3002	Implantable synthetic polymer	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem	29/11/2022	29/11/2022
									A070504	Premature Discharge of Battery	29/11/2022	29/11/2022
							F08	Hospitalization or Prolonged Hospitalization	A0705	Battery Problem	29/11/2022	29/11/2022
									A070504	Premature Discharge of Battery	29/11/2022	29/11/2022
							F1905	Device Revision or Replacement	A0705	Battery Problem	29/11/2022	29/11/2022
									A070504	Premature Discharge of Battery	29/11/2022	29/11/2022
					CT3006	Implantable inorganic material	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem	29/11/2022	29/11/2022
									A070504	Premature Discharge of Battery	29/11/2022	29/11/2022
							F08	Hospitalization or Prolonged Hospitalization	A0705	Battery Problem	29/11/2022	29/11/2022
									A070504	Premature Discharge of Battery	29/11/2022	29/11/2022
							F1905	Device Revision or Replacement	A0705	Battery Problem	29/11/2022	29/11/2022
									A070504	Premature Discharge of Battery	29/11/2022	29/11/2022
					CT314	Electrically powered	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem	29/11/2022	29/11/2022
									A070504	Premature Discharge of Battery	29/11/2022	29/11/2022
							F08	Hospitalization or Prolonged Hospitalization	A0705	Battery Problem	29/11/2022	29/11/2022
									A070504	Premature Discharge of Battery	29/11/2022	29/11/2022
							F1905	Device Revision or Replacement	A0705	Battery Problem	29/11/2022	29/11/2022
									A070504	Premature Discharge of Battery	29/11/2022	29/11/2022
					CT321	Long-term surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem	29/11/2022	29/11/2022
									A070504	Premature Discharge of Battery	29/11/2022	29/11/2022
							F08	Hospitalization or Prolonged Hospitalization	A0705	Battery Problem	29/11/2022	29/11/2022
									A070504	Premature Discharge of Battery	29/11/2022	29/11/2022
							F1905	Device Revision or Replacement	A0705	Battery Problem	29/11/2022	29/11/2022
									A070504	Premature Discharge of Battery	29/11/2022	29/11/2022
					CT449	Generators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem	29/11/2022	29/11/2022
									A070504	Premature Discharge of Battery	29/11/2022	29/11/2022
							F08	Hospitalization or Prolonged Hospitalization	A0705	Battery Problem	29/11/2022	29/11/2022
									A070504	Premature Discharge of Battery	29/11/2022	29/11/2022
							F1905	Device Revision or Replacement	A0705	Battery Problem	29/11/2022	29/11/2022
									A070504	Premature Discharge of Battery	29/11/2022	29/11/2022
					CT981	Single use	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem	29/11/2022	29/11/2022
									A070504	Premature Discharge of Battery	29/11/2022	29/11/2022
							F08	Hospitalization or Prolonged Hospitalization	A0705	Battery Problem	29/11/2022	29/11/2022
									A070504	Premature Discharge of Battery	29/11/2022	29/11/2022
							F1905	Device Revision or Replacement	A0705	Battery Problem	29/11/2022	29/11/2022
									A070504	Premature Discharge of Battery	29/11/2022	29/11/2022
					CT983	Surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem	29/11/2022	29/11/2022
									A070504	Premature Discharge of Battery	29/11/2022	29/11/2022
							F08	Hospitalization or Prolonged Hospitalization	A0705	Battery Problem	29/11/2022	29/11/2022
									A070504	Premature Discharge of Battery	29/11/2022	29/11/2022
							F1905	Device Revision or Replacement	A0705	Battery Problem	29/11/2022	29/11/2022
									A070504	Premature Discharge of Battery	29/11/2022	29/11/2022
11418	2022	All other reportable incidents	47205	Cardiac pulse generator programmer	CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A090805	Incorrect Measurement		
							F23	Unexpected Medical Intervention	A090805	Incorrect Measurement		
11419	2022	All other reportable incidents	47265	Dual-chamber implantable pacemaker, rate-responsive	CT141	Electrodes/Leads and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT145	Cardiovascular devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT233	Surgical	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT2406	Implantable	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT2407	Active implantable	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT2727	Active implant devices/ancillaries	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT2750	Implantable cardiac electrical stimulators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT2985	Active	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT3001	Implantable metal	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		

					CT3002	Implantable synthetic polymer	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT3006	Implantable inorganic material	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT314	Electrically powered	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT321	Long-term surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT449	Generators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT981	Single use	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT983	Surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
11420	2022	All other reportable incidents	47265	Dual-chamber implantable pacemaker, rate-responsive	CT141	Electrodes/Leads and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem	24/10/2022	24/10/2022
									A160601	Premature Elective Replacement Indicator	24/10/2022	24/10/2022
							F08	Hospitalization or Prolonged Hospitalization	A0705	Battery Problem	24/10/2022	24/10/2022
									A160601	Premature Elective Replacement Indicator	24/10/2022	24/10/2022
							F1905	Device Revision or Replacement	A0705	Battery Problem	24/10/2022	24/10/2022
									A160601	Premature Elective Replacement Indicator	24/10/2022	24/10/2022
					CT145	Cardiovascular devices	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem	24/10/2022	24/10/2022
									A160601	Premature Elective Replacement Indicator	24/10/2022	24/10/2022
							F08	Hospitalization or Prolonged Hospitalization	A0705	Battery Problem	24/10/2022	24/10/2022
									A160601	Premature Elective Replacement Indicator	24/10/2022	24/10/2022
							F1905	Device Revision or Replacement	A0705	Battery Problem	24/10/2022	24/10/2022
									A160601	Premature Elective Replacement Indicator	24/10/2022	24/10/2022
					CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem	24/10/2022	24/10/2022
									A160601	Premature Elective Replacement Indicator	24/10/2022	24/10/2022
							F08	Hospitalization or Prolonged Hospitalization	A0705	Battery Problem	24/10/2022	24/10/2022
									A160601	Premature Elective Replacement Indicator	24/10/2022	24/10/2022
							F1905	Device Revision or Replacement	A0705	Battery Problem	24/10/2022	24/10/2022
									A160601	Premature Elective Replacement Indicator	24/10/2022	24/10/2022
					CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem	24/10/2022	24/10/2022
									A160601	Premature Elective Replacement Indicator	24/10/2022	24/10/2022
							F08	Hospitalization or Prolonged Hospitalization	A0705	Battery Problem	24/10/2022	24/10/2022
									A160601	Premature Elective Replacement Indicator	24/10/2022	24/10/2022
							F1905	Device Revision or Replacement	A0705	Battery Problem	24/10/2022	24/10/2022
									A160601	Premature Elective Replacement Indicator	24/10/2022	24/10/2022
					CT233	Surgical	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem	24/10/2022	24/10/2022
									A160601	Premature Elective Replacement Indicator	24/10/2022	24/10/2022
							F08	Hospitalization or Prolonged Hospitalization	A0705	Battery Problem	24/10/2022	24/10/2022
									A160601	Premature Elective Replacement Indicator	24/10/2022	24/10/2022
							F1905	Device Revision or Replacement	A0705	Battery Problem	24/10/2022	24/10/2022
									A160601	Premature Elective Replacement Indicator	24/10/2022	24/10/2022
					CT2406	Implantable	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem	24/10/2022	24/10/2022
									A160601	Premature Elective Replacement Indicator	24/10/2022	24/10/2022
							F08	Hospitalization or Prolonged Hospitalization	A0705	Battery Problem	24/10/2022	24/10/2022
									A160601	Premature Elective Replacement Indicator	24/10/2022	24/10/2022
							F1905	Device Revision or Replacement	A0705	Battery Problem	24/10/2022	24/10/2022
									A160601	Premature Elective Replacement Indicator	24/10/2022	24/10/2022
					CT2407	Active implantable	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem	24/10/2022	24/10/2022
									A160601	Premature Elective Replacement Indicator	24/10/2022	24/10/2022
							F08	Hospitalization or Prolonged Hospitalization	A0705	Battery Problem	24/10/2022	24/10/2022
									A160601	Premature Elective Replacement Indicator	24/10/2022	24/10/2022
							F1905	Device Revision or Replacement	A0705	Battery Problem	24/10/2022	24/10/2022
									A160601	Premature Elective Replacement Indicator	24/10/2022	24/10/2022
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem	24/10/2022	24/10/2022
									A160601	Premature Elective Replacement Indicator	24/10/2022	24/10/2022
							F08	Hospitalization or Prolonged Hospitalization	A0705	Battery Problem	24/10/2022	24/10/2022
									A160601	Premature Elective Replacement Indicator	24/10/2022	24/10/2022
							F1905	Device Revision or Replacement	A0705	Battery Problem	24/10/2022	24/10/2022
									A160601	Premature Elective Replacement Indicator	24/10/2022	24/10/2022
					CT2727	Active implant devices/ancillaries	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem	24/10/2022	24/10/2022
									A160601	Premature Elective Replacement Indicator	24/10/2022	24/10/2022
							F08	Hospitalization or Prolonged Hospitalization	A0705	Battery Problem	24/10/2022	24/10/2022
									A160601	Premature Elective Replacement Indicator	24/10/2022	24/10/2022
							F1905	Device Revision or Replacement	A0705	Battery Problem	24/10/2022	24/10/2022
									A160601	Premature Elective Replacement Indicator	24/10/2022	24/10/2022
					CT2750	Implantable cardiac electrical stimulators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem	24/10/2022	24/10/2022
									A160601	Premature Elective Replacement Indicator	24/10/2022	24/10/2022
							F08	Hospitalization or Prolonged Hospitalization	A0705	Battery Problem	24/10/2022	24/10/2022
									A160601	Premature Elective Replacement Indicator	24/10/2022	24/10/2022
							F1905	Device Revision or Replacement	A0705	Battery Problem	24/10/2022	24/10/2022
									A160601	Premature Elective Replacement Indicator	24/10/2022	24/10/2022
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem	24/10/2022	24/10/2022
									A160601	Premature Elective Replacement Indicator	24/10/2022	24/10/2022
							F08	Hospitalization or Prolonged Hospitalization	A0705	Battery Problem	24/10/2022	24/10/2022
									A160601	Premature Elective Replacement Indicator	24/10/2022	24/10/2022
							F1905	Device Revision or Replacement	A0705	Battery Problem	24/10/2022	24/10/2022
									A160601	Premature Elective Replacement Indicator	24/10/2022	24/10/2022
					CT2985	Active	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem	24/10/2022	24/10/2022
									A160601	Premature Elective Replacement Indicator	24/10/2022	24/10/2022
							F08	Hospitalization or Prolonged Hospitalization	A0705	Battery Problem	24/10/2022	24/10/2022
									A160601	Premature Elective Replacement Indicator	24/10/2022	24/10/2022
							F1905	Device Revision or Replacement	A0705	Battery Problem	24/10/2022	24/10/2022
									A160601	Premature Elective Replacement Indicator	24/10/2022	24/10/2022
					CT3001	Implantable metal	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem	24/10/2022	24/10/2022
									A160601	Premature Elective Replacement Indicator	24/10/2022	24/10/2022

							F08	Hospitalization or Prolonged Hospitalization	A0705	Battery Problem	24/10/2022	24/10/2022
									A160601	Premature Elective Replacement Indicator	24/10/2022	24/10/2022
							F1905	Device Revision or Replacement	A0705	Battery Problem	24/10/2022	24/10/2022
									A160601	Premature Elective Replacement Indicator	24/10/2022	24/10/2022
					CT3002	Implantable synthetic polymer	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem	24/10/2022	24/10/2022
									A160601	Premature Elective Replacement Indicator	24/10/2022	24/10/2022
							F08	Hospitalization or Prolonged Hospitalization	A0705	Battery Problem	24/10/2022	24/10/2022
									A160601	Premature Elective Replacement Indicator	24/10/2022	24/10/2022
							F1905	Device Revision or Replacement	A0705	Battery Problem	24/10/2022	24/10/2022
									A160601	Premature Elective Replacement Indicator	24/10/2022	24/10/2022
					CT3006	Implantable inorganic material	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem	24/10/2022	24/10/2022
									A160601	Premature Elective Replacement Indicator	24/10/2022	24/10/2022
							F08	Hospitalization or Prolonged Hospitalization	A0705	Battery Problem	24/10/2022	24/10/2022
									A160601	Premature Elective Replacement Indicator	24/10/2022	24/10/2022
							F1905	Device Revision or Replacement	A0705	Battery Problem	24/10/2022	24/10/2022
									A160601	Premature Elective Replacement Indicator	24/10/2022	24/10/2022
					CT314	Electrically powered	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem	24/10/2022	24/10/2022
									A160601	Premature Elective Replacement Indicator	24/10/2022	24/10/2022
							F08	Hospitalization or Prolonged Hospitalization	A0705	Battery Problem	24/10/2022	24/10/2022
									A160601	Premature Elective Replacement Indicator	24/10/2022	24/10/2022
							F1905	Device Revision or Replacement	A0705	Battery Problem	24/10/2022	24/10/2022
									A160601	Premature Elective Replacement Indicator	24/10/2022	24/10/2022
					CT321	Long-term surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem	24/10/2022	24/10/2022
									A160601	Premature Elective Replacement Indicator	24/10/2022	24/10/2022
							F08	Hospitalization or Prolonged Hospitalization	A0705	Battery Problem	24/10/2022	24/10/2022
									A160601	Premature Elective Replacement Indicator	24/10/2022	24/10/2022
							F1905	Device Revision or Replacement	A0705	Battery Problem	24/10/2022	24/10/2022
									A160601	Premature Elective Replacement Indicator	24/10/2022	24/10/2022
					CT449	Generators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem	24/10/2022	24/10/2022
									A160601	Premature Elective Replacement Indicator	24/10/2022	24/10/2022
							F08	Hospitalization or Prolonged Hospitalization	A0705	Battery Problem	24/10/2022	24/10/2022
									A160601	Premature Elective Replacement Indicator	24/10/2022	24/10/2022
							F1905	Device Revision or Replacement	A0705	Battery Problem	24/10/2022	24/10/2022
									A160601	Premature Elective Replacement Indicator	24/10/2022	24/10/2022
					CT981	Single use	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem	24/10/2022	24/10/2022
									A160601	Premature Elective Replacement Indicator	24/10/2022	24/10/2022
							F08	Hospitalization or Prolonged Hospitalization	A0705	Battery Problem	24/10/2022	24/10/2022
									A160601	Premature Elective Replacement Indicator	24/10/2022	24/10/2022
							F1905	Device Revision or Replacement	A0705	Battery Problem	24/10/2022	24/10/2022
									A160601	Premature Elective Replacement Indicator	24/10/2022	24/10/2022
					CT983	Surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem	24/10/2022	24/10/2022
									A160601	Premature Elective Replacement Indicator	24/10/2022	24/10/2022
							F08	Hospitalization or Prolonged Hospitalization	A0705	Battery Problem	24/10/2022	24/10/2022
									A160601	Premature Elective Replacement Indicator	24/10/2022	24/10/2022
							F1905	Device Revision or Replacement	A0705	Battery Problem	24/10/2022	24/10/2022
									A160601	Premature Elective Replacement Indicator	24/10/2022	24/10/2022
11421	2022	All other reportable incidents			CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem		
							F26	No Health Consequences or Impact	A0705	Battery Problem		
11422	2022		47265	Dual-chamber implantable pacemaker, rate-responsive	CT141	Electrodes/Leads and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	05/10/2022	05/10/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	05/10/2022	05/10/2022
					CT145	Cardiovascular devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	05/10/2022	05/10/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	05/10/2022	05/10/2022
					CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	05/10/2022	05/10/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	05/10/2022	05/10/2022
					CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	05/10/2022	05/10/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	05/10/2022	05/10/2022
					CT233	Surgical	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	05/10/2022	05/10/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	05/10/2022	05/10/2022
					CT2406	Implantable	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	05/10/2022	05/10/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	05/10/2022	05/10/2022
					CT2407	Active implantable	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	05/10/2022	05/10/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	05/10/2022	05/10/2022
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	05/10/2022	05/10/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	05/10/2022	05/10/2022
					CT2727	Active implant devices/ancillaries	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	05/10/2022	05/10/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	05/10/2022	05/10/2022
					CT2750	Implantable cardiac electrical stimulators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	05/10/2022	05/10/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	05/10/2022	05/10/2022
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	05/10/2022	05/10/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	05/10/2022	05/10/2022
					CT2985	Active	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	05/10/2022	05/10/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	05/10/2022	05/10/2022
					CT3001	Implantable metal	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	05/10/2022	05/10/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	05/10/2022	05/10/2022
					CT3002	Implantable synthetic polymer	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	05/10/2022	05/10/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	05/10/2022	05/10/2022
					CT3006	Implantable inorganic material	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	05/10/2022	05/10/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	05/10/2022	05/10/2022
					CT314	Electrically powered	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	05/10/2022	05/10/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	05/10/2022	05/10/2022
					CT321	Long-term surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	05/10/2022	05/10/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	05/10/2022	05/10/2022
					CT449	Generators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	05/10/2022	05/10/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	05/10/2022	05/10/2022
					CT981	Single use	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	05/10/2022	05/10/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	05/10/2022	05/10/2022
					CT983	Surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	05/10/2022	05/10/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	05/10/2022	05/10/2022

11423	2022	All other reportable incidents	47265	Dual-chamber implantable pacemaker, rate-responsive	CT141	Electrodes/Leads and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A071102	Failure to Interrogate	10/10/2022	10/10/2022
									A071204	No Pacing	10/10/2022	10/10/2022
							F1901	Additional Surgery	A071102	Failure to Interrogate	10/10/2022	10/10/2022
									A071204	No Pacing	10/10/2022	10/10/2022
					CT145	Cardiovascular devices	E2403	No Clinical Signs, Symptoms or Conditions	A071102	Failure to Interrogate	10/10/2022	10/10/2022
									A071204	No Pacing	10/10/2022	10/10/2022
							F1901	Additional Surgery	A071102	Failure to Interrogate	10/10/2022	10/10/2022
									A071204	No Pacing	10/10/2022	10/10/2022
					CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A071102	Failure to Interrogate	10/10/2022	10/10/2022
									A071204	No Pacing	10/10/2022	10/10/2022
							F1901	Additional Surgery	A071102	Failure to Interrogate	10/10/2022	10/10/2022
									A071204	No Pacing	10/10/2022	10/10/2022
					CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A071102	Failure to Interrogate	10/10/2022	10/10/2022
									A071204	No Pacing	10/10/2022	10/10/2022
							F1901	Additional Surgery	A071102	Failure to Interrogate	10/10/2022	10/10/2022
									A071204	No Pacing	10/10/2022	10/10/2022
					CT233	Surgical	E2403	No Clinical Signs, Symptoms or Conditions	A071102	Failure to Interrogate	10/10/2022	10/10/2022
									A071204	No Pacing	10/10/2022	10/10/2022
							F1901	Additional Surgery	A071102	Failure to Interrogate	10/10/2022	10/10/2022
									A071204	No Pacing	10/10/2022	10/10/2022
					CT2406	Implantable	E2403	No Clinical Signs, Symptoms or Conditions	A071102	Failure to Interrogate	10/10/2022	10/10/2022
									A071204	No Pacing	10/10/2022	10/10/2022
							F1901	Additional Surgery	A071102	Failure to Interrogate	10/10/2022	10/10/2022
									A071204	No Pacing	10/10/2022	10/10/2022
					CT2407	Active implantable	E2403	No Clinical Signs, Symptoms or Conditions	A071102	Failure to Interrogate	10/10/2022	10/10/2022
									A071204	No Pacing	10/10/2022	10/10/2022
							F1901	Additional Surgery	A071102	Failure to Interrogate	10/10/2022	10/10/2022
									A071204	No Pacing	10/10/2022	10/10/2022
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A071102	Failure to Interrogate	10/10/2022	10/10/2022
									A071204	No Pacing	10/10/2022	10/10/2022
							F1901	Additional Surgery	A071102	Failure to Interrogate	10/10/2022	10/10/2022
									A071204	No Pacing	10/10/2022	10/10/2022
					CT2727	Active implant devices/ancillaries	E2403	No Clinical Signs, Symptoms or Conditions	A071102	Failure to Interrogate	10/10/2022	10/10/2022
									A071204	No Pacing	10/10/2022	10/10/2022
							F1901	Additional Surgery	A071102	Failure to Interrogate	10/10/2022	10/10/2022
									A071204	No Pacing	10/10/2022	10/10/2022
					CT2750	Implantable cardiac electrical stimulators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A071102	Failure to Interrogate	10/10/2022	10/10/2022
									A071204	No Pacing	10/10/2022	10/10/2022
							F1901	Additional Surgery	A071102	Failure to Interrogate	10/10/2022	10/10/2022
									A071204	No Pacing	10/10/2022	10/10/2022
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A071102	Failure to Interrogate	10/10/2022	10/10/2022
									A071204	No Pacing	10/10/2022	10/10/2022
							F1901	Additional Surgery	A071102	Failure to Interrogate	10/10/2022	10/10/2022
									A071204	No Pacing	10/10/2022	10/10/2022
					CT2985	Active	E2403	No Clinical Signs, Symptoms or Conditions	A071102	Failure to Interrogate	10/10/2022	10/10/2022
									A071204	No Pacing	10/10/2022	10/10/2022
							F1901	Additional Surgery	A071102	Failure to Interrogate	10/10/2022	10/10/2022
									A071204	No Pacing	10/10/2022	10/10/2022
					CT3001	Implantable metal	E2403	No Clinical Signs, Symptoms or Conditions	A071102	Failure to Interrogate	10/10/2022	10/10/2022
									A071204	No Pacing	10/10/2022	10/10/2022
							F1901	Additional Surgery	A071102	Failure to Interrogate	10/10/2022	10/10/2022
									A071204	No Pacing	10/10/2022	10/10/2022
					CT3002	Implantable synthetic polymer	E2403	No Clinical Signs, Symptoms or Conditions	A071102	Failure to Interrogate	10/10/2022	10/10/2022
									A071204	No Pacing	10/10/2022	10/10/2022
							F1901	Additional Surgery	A071102	Failure to Interrogate	10/10/2022	10/10/2022
									A071204	No Pacing	10/10/2022	10/10/2022
					CT3006	Implantable inorganic material	E2403	No Clinical Signs, Symptoms or Conditions	A071102	Failure to Interrogate	10/10/2022	10/10/2022
									A071204	No Pacing	10/10/2022	10/10/2022
							F1901	Additional Surgery	A071102	Failure to Interrogate	10/10/2022	10/10/2022
									A071204	No Pacing	10/10/2022	10/10/2022
					CT314	Electrically powered	E2403	No Clinical Signs, Symptoms or Conditions	A071102	Failure to Interrogate	10/10/2022	10/10/2022
									A071204	No Pacing	10/10/2022	10/10/2022
							F1901	Additional Surgery	A071102	Failure to Interrogate	10/10/2022	10/10/2022
									A071204	No Pacing	10/10/2022	10/10/2022
					CT321	Long-term surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A071102	Failure to Interrogate	10/10/2022	10/10/2022
									A071204	No Pacing	10/10/2022	10/10/2022
							F1901	Additional Surgery	A071102	Failure to Interrogate	10/10/2022	10/10/2022
									A071204	No Pacing	10/10/2022	10/10/2022
					CT449	Generators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A071102	Failure to Interrogate	10/10/2022	10/10/2022
									A071204	No Pacing	10/10/2022	10/10/2022
							F1901	Additional Surgery	A071102	Failure to Interrogate	10/10/2022	10/10/2022
									A071204	No Pacing	10/10/2022	10/10/2022
					CT981	Single use	E2403	No Clinical Signs, Symptoms or Conditions	A071102	Failure to Interrogate	10/10/2022	10/10/2022
									A071204	No Pacing	10/10/2022	10/10/2022
							F1901	Additional Surgery	A071102	Failure to Interrogate	10/10/2022	10/10/2022
									A071204	No Pacing	10/10/2022	10/10/2022
					CT983	Surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A071102	Failure to Interrogate	10/10/2022	10/10/2022
									A071204	No Pacing	10/10/2022	10/10/2022
							F1901	Additional Surgery	A071102	Failure to Interrogate	10/10/2022	10/10/2022
									A071204	No Pacing	10/10/2022	10/10/2022
11424	2022	All other reportable incidents	47265	Dual-chamber implantable pacemaker, rate-responsive	CT141	Electrodes/Leads and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	05/10/2022	05/10/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	05/10/2022	05/10/2022
					CT145	Cardiovascular devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	05/10/2022	05/10/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	05/10/2022	05/10/2022
					CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	05/10/2022	05/10/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	05/10/2022	05/10/2022
					CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	05/10/2022	05/10/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	05/10/2022	05/10/2022

					CT233	Surgical	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	05/10/2022	05/10/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	05/10/2022	05/10/2022
					CT2406	Implantable	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	05/10/2022	05/10/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	05/10/2022	05/10/2022
					CT2407	Active implantable	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	05/10/2022	05/10/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	05/10/2022	05/10/2022
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	05/10/2022	05/10/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	05/10/2022	05/10/2022
					CT2727	Active implant devices/ancillaries	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	05/10/2022	05/10/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	05/10/2022	05/10/2022
					CT2750	Implantable cardiac electrical stimulators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	05/10/2022	05/10/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	05/10/2022	05/10/2022
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	05/10/2022	05/10/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	05/10/2022	05/10/2022
					CT2985	Active	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	05/10/2022	05/10/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	05/10/2022	05/10/2022
					CT3001	Implantable metal	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	05/10/2022	05/10/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	05/10/2022	05/10/2022
					CT3002	Implantable synthetic polymer	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	05/10/2022	05/10/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	05/10/2022	05/10/2022
					CT3006	Implantable inorganic material	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	05/10/2022	05/10/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	05/10/2022	05/10/2022
					CT314	Electrically powered	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	05/10/2022	05/10/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	05/10/2022	05/10/2022
					CT321	Long-term surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	05/10/2022	05/10/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	05/10/2022	05/10/2022
					CT449	Generators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	05/10/2022	05/10/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	05/10/2022	05/10/2022
					CT981	Single use	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	05/10/2022	05/10/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	05/10/2022	05/10/2022
					CT983	Surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	05/10/2022	05/10/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	05/10/2022	05/10/2022
11425	2022		47265	Dual-chamber implantable pacemaker, rate-responsive	CT141	Electrodes/Leads and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	01/10/2022	01/10/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	01/10/2022	01/10/2022
					CT145	Cardiovascular devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	01/10/2022	01/10/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	01/10/2022	01/10/2022
					CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	01/10/2022	01/10/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	01/10/2022	01/10/2022
					CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	01/10/2022	01/10/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	01/10/2022	01/10/2022
					CT233	Surgical	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	01/10/2022	01/10/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	01/10/2022	01/10/2022
					CT2406	Implantable	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	01/10/2022	01/10/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	01/10/2022	01/10/2022
					CT2407	Active implantable	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	01/10/2022	01/10/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	01/10/2022	01/10/2022
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	01/10/2022	01/10/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	01/10/2022	01/10/2022
					CT2727	Active implant devices/ancillaries	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	01/10/2022	01/10/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	01/10/2022	01/10/2022
					CT2750	Implantable cardiac electrical stimulators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	01/10/2022	01/10/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	01/10/2022	01/10/2022
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	01/10/2022	01/10/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	01/10/2022	01/10/2022
					CT2985	Active	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	01/10/2022	01/10/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	01/10/2022	01/10/2022
					CT3001	Implantable metal	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	01/10/2022	01/10/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	01/10/2022	01/10/2022
					CT3002	Implantable synthetic polymer	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	01/10/2022	01/10/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	01/10/2022	01/10/2022
					CT3006	Implantable inorganic material	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	01/10/2022	01/10/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	01/10/2022	01/10/2022
					CT314	Electrically powered	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	01/10/2022	01/10/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	01/10/2022	01/10/2022
					CT321	Long-term surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	01/10/2022	01/10/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	01/10/2022	01/10/2022
					CT449	Generators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	01/10/2022	01/10/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	01/10/2022	01/10/2022
					CT981	Single use	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	01/10/2022	01/10/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	01/10/2022	01/10/2022
					CT983	Surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	01/10/2022	01/10/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	01/10/2022	01/10/2022
11426	2022		47265	Dual-chamber implantable pacemaker, rate-responsive	CT141	Electrodes/Leads and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	05/10/2022	05/10/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	05/10/2022	05/10/2022
					CT145	Cardiovascular devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	05/10/2022	05/10/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	05/10/2022	05/10/2022
					CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	05/10/2022	05/10/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	05/10/2022	05/10/2022
					CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	05/10/2022	05/10/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	05/10/2022	05/10/2022
					CT233	Surgical	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	05/10/2022	05/10/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	05/10/2022	05/10/2022
					CT2406	Implantable	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	05/10/2022	05/10/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	05/10/2022	05/10/2022
					CT2407	Active implantable	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	05/10/2022	05/10/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	05/10/2022	05/10/2022
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	05/10/2022	05/10/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	05/10/2022	05/10/2022

					CT2727	Active implant devices/ancillaries	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	05/10/2022	05/10/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	05/10/2022	05/10/2022
					CT2750	Implantable cardiac electrical stimulators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	05/10/2022	05/10/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	05/10/2022	05/10/2022
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	05/10/2022	05/10/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	05/10/2022	05/10/2022
					CT2985	Active	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	05/10/2022	05/10/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	05/10/2022	05/10/2022
					CT3001	Implantable metal	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	05/10/2022	05/10/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	05/10/2022	05/10/2022
					CT3002	Implantable synthetic polymer	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	05/10/2022	05/10/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	05/10/2022	05/10/2022
					CT3006	Implantable inorganic material	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	05/10/2022	05/10/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	05/10/2022	05/10/2022
					CT314	Electrically powered	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	05/10/2022	05/10/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	05/10/2022	05/10/2022
					CT321	Long-term surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	05/10/2022	05/10/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	05/10/2022	05/10/2022
					CT449	Generators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	05/10/2022	05/10/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	05/10/2022	05/10/2022
					CT981	Single use	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	05/10/2022	05/10/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	05/10/2022	05/10/2022
					CT983	Surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	05/10/2022	05/10/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	05/10/2022	05/10/2022
11427	2022	All other reportable incidents	36241	Pacemaker/defibrillator lead	CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
							F24	Insufficient Health Impact Information	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
11428	2022	All other reportable incidents	47265	Dual-chamber implantable pacemaker, rate-responsive	CT141	Electrodes/Leads and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	17/10/2022	17/10/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	17/10/2022	17/10/2022
					CT145	Cardiovascular devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	17/10/2022	17/10/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	17/10/2022	17/10/2022
					CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	17/10/2022	17/10/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	17/10/2022	17/10/2022
					CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	17/10/2022	17/10/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	17/10/2022	17/10/2022
					CT233	Surgical	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	17/10/2022	17/10/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	17/10/2022	17/10/2022
					CT2406	Implantable	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	17/10/2022	17/10/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	17/10/2022	17/10/2022
					CT2407	Active implantable	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	17/10/2022	17/10/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	17/10/2022	17/10/2022
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	17/10/2022	17/10/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	17/10/2022	17/10/2022
					CT2727	Active implant devices/ancillaries	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	17/10/2022	17/10/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	17/10/2022	17/10/2022
					CT2750	Implantable cardiac electrical stimulators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	17/10/2022	17/10/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	17/10/2022	17/10/2022
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	17/10/2022	17/10/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	17/10/2022	17/10/2022
					CT2985	Active	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	17/10/2022	17/10/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	17/10/2022	17/10/2022
					CT3001	Implantable metal	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	17/10/2022	17/10/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	17/10/2022	17/10/2022
					CT3002	Implantable synthetic polymer	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	17/10/2022	17/10/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	17/10/2022	17/10/2022
					CT3006	Implantable inorganic material	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	17/10/2022	17/10/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	17/10/2022	17/10/2022
					CT314	Electrically powered	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	17/10/2022	17/10/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	17/10/2022	17/10/2022
					CT321	Long-term surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	17/10/2022	17/10/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	17/10/2022	17/10/2022
					CT449	Generators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	17/10/2022	17/10/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	17/10/2022	17/10/2022
					CT981	Single use	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	17/10/2022	17/10/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	17/10/2022	17/10/2022
					CT983	Surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	17/10/2022	17/10/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	17/10/2022	17/10/2022
11429	2022		47265	Dual-chamber implantable pacemaker, rate-responsive	CT141	Electrodes/Leads and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	17/10/2022	17/10/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	17/10/2022	17/10/2022
					CT145	Cardiovascular devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	17/10/2022	17/10/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	17/10/2022	17/10/2022
					CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	17/10/2022	17/10/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	17/10/2022	17/10/2022
					CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	17/10/2022	17/10/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	17/10/2022	17/10/2022
					CT233	Surgical	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	17/10/2022	17/10/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	17/10/2022	17/10/2022
					CT2406	Implantable	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	17/10/2022	17/10/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	17/10/2022	17/10/2022
					CT2407	Active implantable	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	17/10/2022	17/10/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	17/10/2022	17/10/2022
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	17/10/2022	17/10/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	17/10/2022	17/10/2022
					CT2727	Active implant devices/ancillaries	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	17/10/2022	17/10/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	17/10/2022	17/10/2022
					CT2750	Implantable cardiac electrical stimulators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	17/10/2022	17/10/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	17/10/2022	17/10/2022

					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	17/10/2022	17/10/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	17/10/2022	17/10/2022
					CT2985	Active	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	17/10/2022	17/10/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	17/10/2022	17/10/2022
					CT3001	Implantable metal	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	17/10/2022	17/10/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	17/10/2022	17/10/2022
					CT3002	Implantable synthetic polymer	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	17/10/2022	17/10/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	17/10/2022	17/10/2022
					CT3006	Implantable inorganic material	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	17/10/2022	17/10/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	17/10/2022	17/10/2022
					CT314	Electrically powered	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	17/10/2022	17/10/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	17/10/2022	17/10/2022
					CT321	Long-term surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	17/10/2022	17/10/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	17/10/2022	17/10/2022
					CT449	Generators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	17/10/2022	17/10/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	17/10/2022	17/10/2022
					CT981	Single use	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	17/10/2022	17/10/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	17/10/2022	17/10/2022
					CT983	Surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	17/10/2022	17/10/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	17/10/2022	17/10/2022
11430	2022	All other reportable incidents	17882	Physiologic-monitoring defibrillation system	CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070803	Failure to Power Up		
							F27	Problem Identified During Non-Clinical Procedure	A070803	Failure to Power Up		
11431	2022	All other reportable incidents	35852	Single-chamber implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E2104	Electric Shock	A0508	Noise, Audible		
									A070909	Oversensing		
									A071302	Inappropriate/Inadequate Shock/Stimulation		
							F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A0508	Noise, Audible		
									A070909	Oversensing		
									A071302	Inappropriate/Inadequate Shock/Stimulation		
11432	2022	All other reportable incidents	47265	Dual-chamber implantable pacemaker, rate-responsive	CT141	Electrodes/Leads and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	06/09/2022	06/09/2022
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	06/09/2022	06/09/2022
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	06/09/2022	06/09/2022
					CT145	Cardiovascular devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	06/09/2022	06/09/2022
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	06/09/2022	06/09/2022
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	06/09/2022	06/09/2022
					CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	06/09/2022	06/09/2022
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	06/09/2022	06/09/2022
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	06/09/2022	06/09/2022
					CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	06/09/2022	06/09/2022
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	06/09/2022	06/09/2022
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	06/09/2022	06/09/2022
					CT233	Surgical	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	06/09/2022	06/09/2022
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	06/09/2022	06/09/2022
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	06/09/2022	06/09/2022
					CT2406	Implantable	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	06/09/2022	06/09/2022
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	06/09/2022	06/09/2022
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	06/09/2022	06/09/2022
					CT2407	Active implantable	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	06/09/2022	06/09/2022
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	06/09/2022	06/09/2022
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	06/09/2022	06/09/2022
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	06/09/2022	06/09/2022
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	06/09/2022	06/09/2022
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	06/09/2022	06/09/2022
					CT2727	Active implant devices/ancillaries	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	06/09/2022	06/09/2022
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	06/09/2022	06/09/2022
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	06/09/2022	06/09/2022
					CT2750	Implantable cardiac electrical stimulators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	06/09/2022	06/09/2022
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	06/09/2022	06/09/2022
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	06/09/2022	06/09/2022
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	06/09/2022	06/09/2022
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	06/09/2022	06/09/2022
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	06/09/2022	06/09/2022
					CT2985	Active	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	06/09/2022	06/09/2022
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	06/09/2022	06/09/2022
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	06/09/2022	06/09/2022
					CT3001	Implantable metal	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	06/09/2022	06/09/2022
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	06/09/2022	06/09/2022
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	06/09/2022	06/09/2022
					CT3002	Implantable synthetic polymer	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	06/09/2022	06/09/2022
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	06/09/2022	06/09/2022
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	06/09/2022	06/09/2022
					CT3006	Implantable inorganic material	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	06/09/2022	06/09/2022
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	06/09/2022	06/09/2022
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	06/09/2022	06/09/2022
					CT314	Electrically powered	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	06/09/2022	06/09/2022
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	06/09/2022	06/09/2022
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	06/09/2022	06/09/2022
					CT321	Long-term surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	06/09/2022	06/09/2022
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	06/09/2022	06/09/2022
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	06/09/2022	06/09/2022
					CT449	Generators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	06/09/2022	06/09/2022
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	06/09/2022	06/09/2022
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	06/09/2022	06/09/2022
					CT981	Single use	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	06/09/2022	06/09/2022
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	06/09/2022	06/09/2022
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	06/09/2022	06/09/2022
					CT983	Surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	06/09/2022	06/09/2022
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	06/09/2022	06/09/2022
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	06/09/2022	06/09/2022

11433	2022	All other reportable incidents	37265	Dual-chamber implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A070910	Undersensing		
									A0712	Pacing Problem		
							F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A070910	Undersensing		
									A0712	Pacing Problem		
11434	2022		47265	Dual-chamber implantable pacemaker, rate-responsive	CT141	Electrodes/Leads and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	20/10/2022	20/10/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	20/10/2022	20/10/2022
					CT145	Cardiovascular devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	20/10/2022	20/10/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	20/10/2022	20/10/2022
					CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	20/10/2022	20/10/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	20/10/2022	20/10/2022
					CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	20/10/2022	20/10/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	20/10/2022	20/10/2022
					CT233	Surgical	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	20/10/2022	20/10/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	20/10/2022	20/10/2022
					CT2406	Implantable	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	20/10/2022	20/10/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	20/10/2022	20/10/2022
					CT2407	Active implantable	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	20/10/2022	20/10/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	20/10/2022	20/10/2022
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	20/10/2022	20/10/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	20/10/2022	20/10/2022
					CT2727	Active implant devices/ancillaries	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	20/10/2022	20/10/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	20/10/2022	20/10/2022
					CT2750	Implantable cardiac electrical stimulators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	20/10/2022	20/10/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	20/10/2022	20/10/2022
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	20/10/2022	20/10/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	20/10/2022	20/10/2022
					CT2985	Active	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	20/10/2022	20/10/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	20/10/2022	20/10/2022
					CT3001	Implantable metal	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	20/10/2022	20/10/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	20/10/2022	20/10/2022
					CT3002	Implantable synthetic polymer	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	20/10/2022	20/10/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	20/10/2022	20/10/2022
					CT3006	Implantable inorganic material	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	20/10/2022	20/10/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	20/10/2022	20/10/2022
					CT314	Electrically powered	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	20/10/2022	20/10/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	20/10/2022	20/10/2022
					CT321	Long-term surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	20/10/2022	20/10/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	20/10/2022	20/10/2022
					CT449	Generators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	20/10/2022	20/10/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	20/10/2022	20/10/2022
					CT981	Single use	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	20/10/2022	20/10/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	20/10/2022	20/10/2022
					CT983	Surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	20/10/2022	20/10/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	20/10/2022	20/10/2022
11435	2022	All other reportable incidents	47265	Dual-chamber implantable pacemaker, rate-responsive	CT141	Electrodes/Leads and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A0701	Capturing Problem		
									A0712	Pacing Problem		
									A071206	Pacing Asynchronously		
							F26	No Health Consequences or Impact	A0701	Capturing Problem		
									A0712	Pacing Problem		
									A071206	Pacing Asynchronously		
					CT145	Cardiovascular devices	E2403	No Clinical Signs, Symptoms or Conditions	A0701	Capturing Problem		
									A0712	Pacing Problem		
									A071206	Pacing Asynchronously		
							F26	No Health Consequences or Impact	A0701	Capturing Problem		
									A0712	Pacing Problem		
									A071206	Pacing Asynchronously		
					CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A0701	Capturing Problem		
									A0712	Pacing Problem		
									A071206	Pacing Asynchronously		
							F26	No Health Consequences or Impact	A0701	Capturing Problem		
									A0712	Pacing Problem		
									A071206	Pacing Asynchronously		
					CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A0701	Capturing Problem		
									A0712	Pacing Problem		
									A071206	Pacing Asynchronously		
							F26	No Health Consequences or Impact	A0701	Capturing Problem		
									A0712	Pacing Problem		
									A071206	Pacing Asynchronously		
					CT233	Surgical	E2403	No Clinical Signs, Symptoms or Conditions	A0701	Capturing Problem		
									A0712	Pacing Problem		
									A071206	Pacing Asynchronously		
							F26	No Health Consequences or Impact	A0701	Capturing Problem		
									A0712	Pacing Problem		
									A071206	Pacing Asynchronously		
					CT2406	Implantable	E2403	No Clinical Signs, Symptoms or Conditions	A0701	Capturing Problem		
									A0712	Pacing Problem		
									A071206	Pacing Asynchronously		
							F26	No Health Consequences or Impact	A0701	Capturing Problem		
									A0712	Pacing Problem		
									A071206	Pacing Asynchronously		
					CT2407	Active implantable	E2403	No Clinical Signs, Symptoms or Conditions	A0701	Capturing Problem		
									A0712	Pacing Problem		
									A071206	Pacing Asynchronously		
							F26	No Health Consequences or Impact	A0701	Capturing Problem		
									A0712	Pacing Problem		
									A071206	Pacing Asynchronously		
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A0701	Capturing Problem		
									A0712	Pacing Problem		

									A071206	Pacing Asynchronously			
							F26	No Health Consequences or Impact	A0701	Capturing Problem			
									A0712	Pacing Problem			
									A071206	Pacing Asynchronously			
					CT2727	Active implant devices/ancillaries	E2403	No Clinical Signs, Symptoms or Conditions	A0701	Capturing Problem			
									A0712	Pacing Problem			
									A071206	Pacing Asynchronously			
							F26	No Health Consequences or Impact	A0701	Capturing Problem			
									A0712	Pacing Problem			
									A071206	Pacing Asynchronously			
					CT2750	Implantable cardiac electrical stimulators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A0701	Capturing Problem			
									A0712	Pacing Problem			
									A071206	Pacing Asynchronously			
							F26	No Health Consequences or Impact	A0701	Capturing Problem			
									A0712	Pacing Problem			
									A071206	Pacing Asynchronously			
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A0701	Capturing Problem			
									A0712	Pacing Problem			
									A071206	Pacing Asynchronously			
							F26	No Health Consequences or Impact	A0701	Capturing Problem			
									A0712	Pacing Problem			
									A071206	Pacing Asynchronously			
					CT2985	Active	E2403	No Clinical Signs, Symptoms or Conditions	A0701	Capturing Problem			
									A0712	Pacing Problem			
									A071206	Pacing Asynchronously			
							F26	No Health Consequences or Impact	A0701	Capturing Problem			
									A0712	Pacing Problem			
									A071206	Pacing Asynchronously			
					CT3001	Implantable metal	E2403	No Clinical Signs, Symptoms or Conditions	A0701	Capturing Problem			
									A0712	Pacing Problem			
									A071206	Pacing Asynchronously			
							F26	No Health Consequences or Impact	A0701	Capturing Problem			
									A0712	Pacing Problem			
									A071206	Pacing Asynchronously			
					CT3002	Implantable synthetic polymer	E2403	No Clinical Signs, Symptoms or Conditions	A0701	Capturing Problem			
									A0712	Pacing Problem			
									A071206	Pacing Asynchronously			
							F26	No Health Consequences or Impact	A0701	Capturing Problem			
									A0712	Pacing Problem			
									A071206	Pacing Asynchronously			
					CT3006	Implantable inorganic material	E2403	No Clinical Signs, Symptoms or Conditions	A0701	Capturing Problem			
									A0712	Pacing Problem			
									A071206	Pacing Asynchronously			
							F26	No Health Consequences or Impact	A0701	Capturing Problem			
									A0712	Pacing Problem			
									A071206	Pacing Asynchronously			
					CT314	Electrically powered	E2403	No Clinical Signs, Symptoms or Conditions	A0701	Capturing Problem			
									A0712	Pacing Problem			
									A071206	Pacing Asynchronously			
							F26	No Health Consequences or Impact	A0701	Capturing Problem			
									A0712	Pacing Problem			
									A071206	Pacing Asynchronously			
					CT321	Long-term surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A0701	Capturing Problem			
									A0712	Pacing Problem			
									A071206	Pacing Asynchronously			
							F26	No Health Consequences or Impact	A0701	Capturing Problem			
									A0712	Pacing Problem			
									A071206	Pacing Asynchronously			
					CT449	Generators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A0701	Capturing Problem			
									A0712	Pacing Problem			
									A071206	Pacing Asynchronously			
							F26	No Health Consequences or Impact	A0701	Capturing Problem			
									A0712	Pacing Problem			
									A071206	Pacing Asynchronously			
					CT981	Single use	E2403	No Clinical Signs, Symptoms or Conditions	A0701	Capturing Problem			
									A0712	Pacing Problem			
									A071206	Pacing Asynchronously			
							F26	No Health Consequences or Impact	A0701	Capturing Problem			
									A0712	Pacing Problem			
									A071206	Pacing Asynchronously			
					CT983	Surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A0701	Capturing Problem			
									A0712	Pacing Problem			
									A071206	Pacing Asynchronously			
							F26	No Health Consequences or Impact	A0701	Capturing Problem			
									A0712	Pacing Problem			
									A071206	Pacing Asynchronously			
11436	2022		47265	Dual-chamber implantable pacemaker, rate-responsive	CT141	Electrodes/Leads and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	18/10/2022	18/10/2022	
							F1901	Additional Surgery	A24	No Apparent Device Problem	18/10/2022	18/10/2022	
					CT145	Cardiovascular devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	18/10/2022	18/10/2022	
							F1901	Additional Surgery	A24	No Apparent Device Problem	18/10/2022	18/10/2022	
					CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	18/10/2022	18/10/2022	
							F1901	Additional Surgery	A24	No Apparent Device Problem	18/10/2022	18/10/2022	
					CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	18/10/2022	18/10/2022	
							F1901	Additional Surgery	A24	No Apparent Device Problem	18/10/2022	18/10/2022	
					CT233	Surgical	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	18/10/2022	18/10/2022	
							F1901	Additional Surgery	A24	No Apparent Device Problem	18/10/2022	18/10/2022	
					CT2406	Implantable	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	18/10/2022	18/10/2022	
							F1901	Additional Surgery	A24	No Apparent Device Problem	18/10/2022	18/10/2022	

					CT2407	Active implantable	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	18/10/2022	18/10/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	18/10/2022	18/10/2022
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	18/10/2022	18/10/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	18/10/2022	18/10/2022
					CT2727	Active implant devices/ancillaries	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	18/10/2022	18/10/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	18/10/2022	18/10/2022
					CT2750	Implantable cardiac electrical stimulators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	18/10/2022	18/10/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	18/10/2022	18/10/2022
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	18/10/2022	18/10/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	18/10/2022	18/10/2022
					CT2985	Active	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	18/10/2022	18/10/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	18/10/2022	18/10/2022
					CT3001	Implantable metal	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	18/10/2022	18/10/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	18/10/2022	18/10/2022
					CT3002	Implantable synthetic polymer	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	18/10/2022	18/10/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	18/10/2022	18/10/2022
					CT3006	Implantable inorganic material	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	18/10/2022	18/10/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	18/10/2022	18/10/2022
					CT314	Electrically powered	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	18/10/2022	18/10/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	18/10/2022	18/10/2022
					CT321	Long-term surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	18/10/2022	18/10/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	18/10/2022	18/10/2022
					CT449	Generators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	18/10/2022	18/10/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	18/10/2022	18/10/2022
					CT981	Single use	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	18/10/2022	18/10/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	18/10/2022	18/10/2022
					CT983	Surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	18/10/2022	18/10/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	18/10/2022	18/10/2022
11437	2022		35223	Endocardial/interventricular septal pacing lead	CT1684	Cardiac pacemakers and associated devices	E2311	Discomfort	A24	No Apparent Device Problem		
							F23	Unexpected Medical Intervention	A24	No Apparent Device Problem		
11438	2022	All other reportable incidents	47265	Dual-chamber implantable pacemaker, rate-responsive	CT141	Electrodes/Leads and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A071102	Failure to Interrogate	19/10/2022	19/10/2022
							F1901	Additional Surgery	A071102	Failure to Interrogate	19/10/2022	19/10/2022
					CT145	Cardiovascular devices	E2403	No Clinical Signs, Symptoms or Conditions	A071102	Failure to Interrogate	19/10/2022	19/10/2022
							F1901	Additional Surgery	A071102	Failure to Interrogate	19/10/2022	19/10/2022
					CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A071102	Failure to Interrogate	19/10/2022	19/10/2022
							F1901	Additional Surgery	A071102	Failure to Interrogate	19/10/2022	19/10/2022
					CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A071102	Failure to Interrogate	19/10/2022	19/10/2022
							F1901	Additional Surgery	A071102	Failure to Interrogate	19/10/2022	19/10/2022
					CT233	Surgical	E2403	No Clinical Signs, Symptoms or Conditions	A071102	Failure to Interrogate	19/10/2022	19/10/2022
							F1901	Additional Surgery	A071102	Failure to Interrogate	19/10/2022	19/10/2022
					CT2406	Implantable	E2403	No Clinical Signs, Symptoms or Conditions	A071102	Failure to Interrogate	19/10/2022	19/10/2022
							F1901	Additional Surgery	A071102	Failure to Interrogate	19/10/2022	19/10/2022
					CT2407	Active implantable	E2403	No Clinical Signs, Symptoms or Conditions	A071102	Failure to Interrogate	19/10/2022	19/10/2022
							F1901	Additional Surgery	A071102	Failure to Interrogate	19/10/2022	19/10/2022
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A071102	Failure to Interrogate	19/10/2022	19/10/2022
							F1901	Additional Surgery	A071102	Failure to Interrogate	19/10/2022	19/10/2022
					CT2727	Active implant devices/ancillaries	E2403	No Clinical Signs, Symptoms or Conditions	A071102	Failure to Interrogate	19/10/2022	19/10/2022
							F1901	Additional Surgery	A071102	Failure to Interrogate	19/10/2022	19/10/2022
					CT2750	Implantable cardiac electrical stimulators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A071102	Failure to Interrogate	19/10/2022	19/10/2022
							F1901	Additional Surgery	A071102	Failure to Interrogate	19/10/2022	19/10/2022
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A071102	Failure to Interrogate	19/10/2022	19/10/2022
							F1901	Additional Surgery	A071102	Failure to Interrogate	19/10/2022	19/10/2022
					CT2985	Active	E2403	No Clinical Signs, Symptoms or Conditions	A071102	Failure to Interrogate	19/10/2022	19/10/2022
							F1901	Additional Surgery	A071102	Failure to Interrogate	19/10/2022	19/10/2022
					CT3001	Implantable metal	E2403	No Clinical Signs, Symptoms or Conditions	A071102	Failure to Interrogate	19/10/2022	19/10/2022
							F1901	Additional Surgery	A071102	Failure to Interrogate	19/10/2022	19/10/2022
					CT3002	Implantable synthetic polymer	E2403	No Clinical Signs, Symptoms or Conditions	A071102	Failure to Interrogate	19/10/2022	19/10/2022
							F1901	Additional Surgery	A071102	Failure to Interrogate	19/10/2022	19/10/2022
					CT3006	Implantable inorganic material	E2403	No Clinical Signs, Symptoms or Conditions	A071102	Failure to Interrogate	19/10/2022	19/10/2022
							F1901	Additional Surgery	A071102	Failure to Interrogate	19/10/2022	19/10/2022
					CT314	Electrically powered	E2403	No Clinical Signs, Symptoms or Conditions	A071102	Failure to Interrogate	19/10/2022	19/10/2022
							F1901	Additional Surgery	A071102	Failure to Interrogate	19/10/2022	19/10/2022
					CT321	Long-term surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A071102	Failure to Interrogate	19/10/2022	19/10/2022
							F1901	Additional Surgery	A071102	Failure to Interrogate	19/10/2022	19/10/2022
					CT449	Generators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A071102	Failure to Interrogate	19/10/2022	19/10/2022
							F1901	Additional Surgery	A071102	Failure to Interrogate	19/10/2022	19/10/2022
					CT981	Single use	E2403	No Clinical Signs, Symptoms or Conditions	A071102	Failure to Interrogate	19/10/2022	19/10/2022
							F1901	Additional Surgery	A071102	Failure to Interrogate	19/10/2022	19/10/2022
					CT983	Surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A071102	Failure to Interrogate	19/10/2022	19/10/2022
							F1901	Additional Surgery	A071102	Failure to Interrogate	19/10/2022	19/10/2022
11439	2022		47265	Dual-chamber implantable pacemaker, rate-responsive	CT141	Electrodes/Leads and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	20/10/2022	20/10/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	20/10/2022	20/10/2022
					CT145	Cardiovascular devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	20/10/2022	20/10/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	20/10/2022	20/10/2022
					CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	20/10/2022	20/10/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	20/10/2022	20/10/2022
					CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	20/10/2022	20/10/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	20/10/2022	20/10/2022
					CT233	Surgical	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	20/10/2022	20/10/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	20/10/2022	20/10/2022
					CT2406	Implantable	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	20/10/2022	20/10/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	20/10/2022	20/10/2022
					CT2407	Active implantable	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	20/10/2022	20/10/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	20/10/2022	20/10/2022
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	20/10/2022	20/10/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	20/10/2022	20/10/2022
					CT2727	Active implant devices/ancillaries	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	20/10/2022	20/10/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	20/10/2022	20/10/2022

					CT2750	Implantable cardiac electrical stimulators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	20/10/2022	20/10/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	20/10/2022	20/10/2022
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	20/10/2022	20/10/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	20/10/2022	20/10/2022
					CT2985	Active	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	20/10/2022	20/10/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	20/10/2022	20/10/2022
					CT3001	Implantable metal	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	20/10/2022	20/10/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	20/10/2022	20/10/2022
					CT3002	Implantable synthetic polymer	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	20/10/2022	20/10/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	20/10/2022	20/10/2022
					CT3006	Implantable inorganic material	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	20/10/2022	20/10/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	20/10/2022	20/10/2022
					CT314	Electrically powered	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	20/10/2022	20/10/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	20/10/2022	20/10/2022
					CT321	Long-term surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	20/10/2022	20/10/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	20/10/2022	20/10/2022
					CT449	Generators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	20/10/2022	20/10/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	20/10/2022	20/10/2022
					CT981	Single use	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	20/10/2022	20/10/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	20/10/2022	20/10/2022
					CT983	Surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	20/10/2022	20/10/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	20/10/2022	20/10/2022
11440	2022	All other reportable incidents	35852	Single-chamber implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem	20/10/2022	20/10/2022
									A070504	Premature Discharge of Battery	20/10/2022	20/10/2022
							F08	Hospitalization or Prolonged Hospitalization	A0705	Battery Problem	20/10/2022	20/10/2022
									A070504	Premature Discharge of Battery	20/10/2022	20/10/2022
							F1905	Device Revision or Replacement	A0705	Battery Problem	20/10/2022	20/10/2022
									A070504	Premature Discharge of Battery	20/10/2022	20/10/2022
11441	2022	All other reportable incidents	37265	Dual-chamber implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A070909	Oversensing		
									A0712	Pacing Problem		
							F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A070909	Oversensing		
									A0712	Pacing Problem		
11442	2022	All other reportable incidents	47265	Dual-chamber implantable pacemaker, rate-responsive	CT141	Electrodes/Leads and associated devices	E060104	Bradycardia	A070101	Failure to Capture		
							F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A070101	Failure to Capture		
					CT145	Cardiovascular devices	E060104	Bradycardia	A070101	Failure to Capture		
							F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A070101	Failure to Capture		
					CT1684	Cardiac pacemakers and associated devices	E060104	Bradycardia	A070101	Failure to Capture		
							F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A070101	Failure to Capture		
					CT1685	Cardiac pacemakers	E060104	Bradycardia	A070101	Failure to Capture		
							F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A070101	Failure to Capture		
					CT233	Surgical	E060104	Bradycardia	A070101	Failure to Capture		
							F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A070101	Failure to Capture		
					CT2406	Implantable	E060104	Bradycardia	A070101	Failure to Capture		
							F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A070101	Failure to Capture		
					CT2407	Active implantable	E060104	Bradycardia	A070101	Failure to Capture		
							F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A070101	Failure to Capture		
					CT257	Implantable cardiac pacemakers	E060104	Bradycardia	A070101	Failure to Capture		
							F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A070101	Failure to Capture		
					CT2727	Active implant devices/ancillaries	E060104	Bradycardia	A070101	Failure to Capture		
							F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A070101	Failure to Capture		
					CT2750	Implantable cardiac electrical stimulators and associated devices	E060104	Bradycardia	A070101	Failure to Capture		
							F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A070101	Failure to Capture		
					CT2866	Implantable cardiac electrical stimulators	E060104	Bradycardia	A070101	Failure to Capture		
							F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A070101	Failure to Capture		
					CT2985	Active	E060104	Bradycardia	A070101	Failure to Capture		
							F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A070101	Failure to Capture		
					CT3001	Implantable metal	E060104	Bradycardia	A070101	Failure to Capture		
							F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A070101	Failure to Capture		
					CT3002	Implantable synthetic polymer	E060104	Bradycardia	A070101	Failure to Capture		
							F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A070101	Failure to Capture		
					CT3006	Implantable inorganic material	E060104	Bradycardia	A070101	Failure to Capture		
							F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A070101	Failure to Capture		
					CT314	Electrically powered	E060104	Bradycardia	A070101	Failure to Capture		
							F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A070101	Failure to Capture		
					CT321	Long-term surgical invasive	E060104	Bradycardia	A070101	Failure to Capture		
							F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A070101	Failure to Capture		
					CT449	Generators and associated devices	E060104	Bradycardia	A070101	Failure to Capture		
							F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A070101	Failure to Capture		
					CT981	Single use	E060104	Bradycardia	A070101	Failure to Capture		
							F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A070101	Failure to Capture		
					CT983	Surgical invasive	E060104	Bradycardia	A070101	Failure to Capture		
							F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A070101	Failure to Capture		
11443	2022	All other reportable incidents	47263	Cardiac resynchronization therapy implantable pacemaker	CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	12/10/2022	12/10/2022
									A0908	Incorrect, Inadequate or Imprecise Result or Readings	12/10/2022	12/10/2022
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	12/10/2022	12/10/2022
									A0908	Incorrect, Inadequate or Imprecise Result or Readings	12/10/2022	12/10/2022
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	12/10/2022	12/10/2022
									A0908	Incorrect, Inadequate or Imprecise Result or Readings	12/10/2022	12/10/2022
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	12/10/2022	12/10/2022
									A0908	Incorrect, Inadequate or Imprecise Result or Readings	12/10/2022	12/10/2022
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	12/10/2022	12/10/2022
									A0908	Incorrect, Inadequate or Imprecise Result or Readings	12/10/2022	12/10/2022
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	12/10/2022	12/10/2022

									A0908	Incorrect, Inadequate or Imprecise Result or Readings	12/10/2022	12/10/2022
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	12/10/2022	12/10/2022
									A0908	Incorrect, Inadequate or Imprecise Result or Readings	12/10/2022	12/10/2022
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	12/10/2022	12/10/2022
									A0908	Incorrect, Inadequate or Imprecise Result or Readings	12/10/2022	12/10/2022
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	12/10/2022	12/10/2022
									A0908	Incorrect, Inadequate or Imprecise Result or Readings	12/10/2022	12/10/2022
11444	2022	All other reportable incidents	47263	Cardiac resynchronization therapy implantable pacemaker	CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	27/10/2022	27/10/2022
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	27/10/2022	27/10/2022
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	27/10/2022	27/10/2022
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	27/10/2022	27/10/2022
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	27/10/2022	27/10/2022
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	27/10/2022	27/10/2022
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	27/10/2022	27/10/2022
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	27/10/2022	27/10/2022
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	27/10/2022	27/10/2022
11445	2022		47265	Dual-chamber implantable pacemaker, rate-responsive	CT141	Electrodes/Leads and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	10/10/2022	10/10/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	10/10/2022	10/10/2022
					CT145	Cardiovascular devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	10/10/2022	10/10/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	10/10/2022	10/10/2022
					CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	10/10/2022	10/10/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	10/10/2022	10/10/2022
					CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	10/10/2022	10/10/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	10/10/2022	10/10/2022
					CT233	Surgical	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	10/10/2022	10/10/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	10/10/2022	10/10/2022
					CT2406	Implantable	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	10/10/2022	10/10/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	10/10/2022	10/10/2022
					CT2407	Active implantable	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	10/10/2022	10/10/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	10/10/2022	10/10/2022
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	10/10/2022	10/10/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	10/10/2022	10/10/2022
					CT2727	Active implant devices/ancillaries	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	10/10/2022	10/10/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	10/10/2022	10/10/2022
					CT2750	Implantable cardiac electrical stimulators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	10/10/2022	10/10/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	10/10/2022	10/10/2022
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	10/10/2022	10/10/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	10/10/2022	10/10/2022
					CT2985	Active	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	10/10/2022	10/10/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	10/10/2022	10/10/2022
					CT3001	Implantable metal	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	10/10/2022	10/10/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	10/10/2022	10/10/2022
					CT3002	Implantable synthetic polymer	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	10/10/2022	10/10/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	10/10/2022	10/10/2022
					CT3006	Implantable inorganic material	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	10/10/2022	10/10/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	10/10/2022	10/10/2022
					CT314	Electrically powered	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	10/10/2022	10/10/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	10/10/2022	10/10/2022
					CT321	Long-term surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	10/10/2022	10/10/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	10/10/2022	10/10/2022
					CT449	Generators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	10/10/2022	10/10/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	10/10/2022	10/10/2022
					CT981	Single use	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	10/10/2022	10/10/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	10/10/2022	10/10/2022
					CT983	Surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	10/10/2022	10/10/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	10/10/2022	10/10/2022
11446	2022	Unanticipated serious deterioration in state of health	60789	Intracardiac pacemaker	CT130	Catheters and associated devices						
					CT145	Cardiovascular devices						
					CT1684	Cardiac pacemakers and associated devices						
					CT1685	Cardiac pacemakers						
					CT233	Surgical						
					CT2406	Implantable						
					CT2407	Active implantable						
					CT257	Implantable cardiac pacemakers						
					CT2727	Active implant devices/ancillaries						
					CT2750	Implantable cardiac electrical stimulators and associated devices						
					CT2818	Catheterization/Cannulation/Fluid path devices						
					CT2866	Implantable cardiac electrical stimulators						
					CT2985	Active						
					CT3001	Implantable metal						
					CT3002	Implantable synthetic polymer						
					CT3006	Implantable inorganic material						
					CT314	Electrically powered						
					CT321	Long-term surgical invasive						
					CT981	Single use						
					CT983	Surgical invasive						
11447	2022	All other reportable incidents	17882	Physiologic-monitoring defibrillation system	CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070803	Failure to Power Up		
							F27	Problem Identified During Non-Clinical Procedure	A070803	Failure to Power Up		
11448	2022	All other reportable incidents	35852	Single-chamber implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	04/11/2022	04/11/2022
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	04/11/2022	04/11/2022
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	04/11/2022	04/11/2022
11449	2022	All other reportable incidents	17882	Physiologic-monitoring defibrillation system	CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070803	Failure to Power Up		
							F27	Problem Identified During Non-Clinical Procedure	A070803	Failure to Power Up		
11450	2022	All other reportable incidents	47265	Dual-chamber implantable pacemaker, rate-	CT141	Electrodes/Leads and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		

				responsive			F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT145	Cardiovascular devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT233	Surgical	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT2406	Implantable	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT2407	Active implantable	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT2727	Active implant devices/ancillaries	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT2750	Implantable cardiac electrical stimulators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT2985	Active	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT3001	Implantable metal	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT3002	Implantable synthetic polymer	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT3006	Implantable inorganic material	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT314	Electrically powered	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT321	Long-term surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT449	Generators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT981	Single use	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT983	Surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
11451	2022	All other reportable incidents	60789	Intracardiac pacemaker	CT130	Catheters and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A26	Insufficient Device Problem Information		
					CT145	Cardiovascular devices	E2403	No Clinical Signs, Symptoms or Conditions	A26	Insufficient Device Problem Information		
					CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A26	Insufficient Device Problem Information		
					CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A26	Insufficient Device Problem Information		
					CT233	Surgical	E2403	No Clinical Signs, Symptoms or Conditions	A26	Insufficient Device Problem Information		
					CT2406	Implantable	E2403	No Clinical Signs, Symptoms or Conditions	A26	Insufficient Device Problem Information		
					CT2407	Active implantable	E2403	No Clinical Signs, Symptoms or Conditions	A26	Insufficient Device Problem Information		
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A26	Insufficient Device Problem Information		
					CT2727	Active implant devices/ancillaries	E2403	No Clinical Signs, Symptoms or Conditions	A26	Insufficient Device Problem Information		
					CT2750	Implantable cardiac electrical stimulators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A26	Insufficient Device Problem Information		
					CT2818	Catheterization/Cannulation/Fluid path devices	E2403	No Clinical Signs, Symptoms or Conditions	A26	Insufficient Device Problem Information		
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A26	Insufficient Device Problem Information		
					CT2985	Active	E2403	No Clinical Signs, Symptoms or Conditions	A26	Insufficient Device Problem Information		
					CT3001	Implantable metal	E2403	No Clinical Signs, Symptoms or Conditions	A26	Insufficient Device Problem Information		
					CT3002	Implantable synthetic polymer	E2403	No Clinical Signs, Symptoms or Conditions	A26	Insufficient Device Problem Information		
					CT3006	Implantable inorganic material	E2403	No Clinical Signs, Symptoms or Conditions	A26	Insufficient Device Problem Information		
					CT314	Electrically powered	E2403	No Clinical Signs, Symptoms or Conditions	A26	Insufficient Device Problem Information		
					CT321	Long-term surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A26	Insufficient Device Problem Information		
					CT981	Single use	E2403	No Clinical Signs, Symptoms or Conditions	A26	Insufficient Device Problem Information		
					CT983	Surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A26	Insufficient Device Problem Information		
11452	2022	All other reportable incidents	37265	Dual-chamber implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A070909	Oversensing		
							F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A070909	Oversensing		
11453	2022	All other reportable incidents	35852	Single-chamber implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem	08/02/2023	08/02/2023
									A070504	Premature Discharge of Battery	08/02/2023	08/02/2023
							F08	Hospitalization or Prolonged Hospitalization	A0705	Battery Problem	08/02/2023	08/02/2023
									A070504	Premature Discharge of Battery	08/02/2023	08/02/2023
							F1905	Device Revision or Replacement	A0705	Battery Problem	08/02/2023	08/02/2023
									A070504	Premature Discharge of Battery	08/02/2023	08/02/2023
11454	2022	All other reportable incidents	47265	Dual-chamber implantable pacemaker, rate-responsive	CT141	Electrodes/Leads and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT145	Cardiovascular devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT233	Surgical	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT2406	Implantable	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT2407	Active implantable	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT2727	Active implant devices/ancillaries	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT2750	Implantable cardiac electrical stimulators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		

					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT2985	Active	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT3001	Implantable metal	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT3002	Implantable synthetic polymer	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT3006	Implantable inorganic material	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT314	Electrically powered	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT321	Long-term surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT449	Generators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT981	Single use	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT983	Surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
11455	2022	All other reportable incidents	35852	Single-chamber implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E2104	Electric Shock	A0709	Device Sensing Problem		
									A070909	Oversensing		
									A071302	Inappropriate/Inadequate Shock/Stimulation		
							F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A0709	Device Sensing Problem		
									A070909	Oversensing		
									A071302	Inappropriate/Inadequate Shock/Stimulation		
							F2203	Imaging Required	A0709	Device Sensing Problem		
									A070909	Oversensing		
									A071302	Inappropriate/Inadequate Shock/Stimulation		
11456	2022	All other reportable incidents	36241	Pacemaker/defibrillator lead	CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A072201	High Impedance		
							F26	No Health Consequences or Impact	A072201	High Impedance		
11457	2022	All other reportable incidents	35852	Single-chamber implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E2104	Electric Shock	A0709	Device Sensing Problem		
									A070909	Oversensing		
									A071302	Inappropriate/Inadequate Shock/Stimulation		
							F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A0709	Device Sensing Problem		
									A070909	Oversensing		
									A071302	Inappropriate/Inadequate Shock/Stimulation		
							F2203	Imaging Required	A0709	Device Sensing Problem		
									A070909	Oversensing		
									A071302	Inappropriate/Inadequate Shock/Stimulation		
11458	2022	All other reportable incidents	47265	Dual-chamber implantable pacemaker, rate-responsive	CT141	Electrodes/Leads and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
					CT145	Cardiovascular devices	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
					CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
					CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
					CT233	Surgical	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
					CT2406	Implantable	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
					CT2407	Active implantable	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
					CT2727	Active implant devices/ancillaries	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
					CT2750	Implantable cardiac electrical stimulators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
					CT2985	Active	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		

					CT3001	Implantable metal	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
					CT3002	Implantable synthetic polymer	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
					CT3006	Implantable inorganic material	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
					CT314	Electrically powered	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
					CT321	Long-term surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
					CT449	Generators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
					CT981	Single use	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
					CT983	Surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
11459	2022	All other reportable incidents	35852	Single-chamber implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E2104	Electric Shock	A070909	Oversensing		
									A071302	Inappropriate/Inadequate Shock/Stimulation		
									A090801	Signal Artifact/Noise		
							F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A070909	Oversensing		
									A071302	Inappropriate/Inadequate Shock/Stimulation		
									A090801	Signal Artifact/Noise		
							F2203	Imaging Required	A070909	Oversensing		
									A071302	Inappropriate/Inadequate Shock/Stimulation		
									A090801	Signal Artifact/Noise		
11460	2022	Unanticipated serious deterioration in state of health	60789	Intracardiac pacemaker	CT130	Catheters and associated devices			A0101	Patient-Device Incompatibility		
					CT145	Cardiovascular devices			A0101	Patient-Device Incompatibility		
					CT1684	Cardiac pacemakers and associated devices			A0101	Patient-Device Incompatibility		
					CT1685	Cardiac pacemakers			A0101	Patient-Device Incompatibility		
					CT233	Surgical			A0101	Patient-Device Incompatibility		
					CT2406	Implantable			A0101	Patient-Device Incompatibility		
					CT2407	Active implantable			A0101	Patient-Device Incompatibility		
					CT257	Implantable cardiac pacemakers			A0101	Patient-Device Incompatibility		
					CT2727	Active implant devices/ancillaries			A0101	Patient-Device Incompatibility		
					CT2750	Implantable cardiac electrical stimulators and associated devices			A0101	Patient-Device Incompatibility		
					CT2818	Catheterization/Cannulation/Fluid path devices			A0101	Patient-Device Incompatibility		
					CT2866	Implantable cardiac electrical stimulators			A0101	Patient-Device Incompatibility		
					CT2985	Active			A0101	Patient-Device Incompatibility		
					CT3001	Implantable metal			A0101	Patient-Device Incompatibility		
					CT3002	Implantable synthetic polymer			A0101	Patient-Device Incompatibility		
					CT3006	Implantable inorganic material			A0101	Patient-Device Incompatibility		
					CT314	Electrically powered			A0101	Patient-Device Incompatibility		
					CT321	Long-term surgical invasive			A0101	Patient-Device Incompatibility		
					CT981	Single use			A0101	Patient-Device Incompatibility		
					CT983	Surgical invasive			A0101	Patient-Device Incompatibility		
11461	2022	All other reportable incidents	47265	Dual-chamber implantable pacemaker, rate-responsive	CT141	Electrodes/Leads and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A071102	Failure to Interrogate	25/10/2022	25/10/2022
							F1901	Additional Surgery	A071102	Failure to Interrogate	25/10/2022	25/10/2022
					CT145	Cardiovascular devices	E2403	No Clinical Signs, Symptoms or Conditions	A071102	Failure to Interrogate	25/10/2022	25/10/2022
							F1901	Additional Surgery	A071102	Failure to Interrogate	25/10/2022	25/10/2022
					CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A071102	Failure to Interrogate	25/10/2022	25/10/2022
							F1901	Additional Surgery	A071102	Failure to Interrogate	25/10/2022	25/10/2022
					CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A071102	Failure to Interrogate	25/10/2022	25/10/2022
							F1901	Additional Surgery	A071102	Failure to Interrogate	25/10/2022	25/10/2022
					CT233	Surgical	E2403	No Clinical Signs, Symptoms or Conditions	A071102	Failure to Interrogate	25/10/2022	25/10/2022
							F1901	Additional Surgery	A071102	Failure to Interrogate	25/10/2022	25/10/2022
					CT2406	Implantable	E2403	No Clinical Signs, Symptoms or Conditions	A071102	Failure to Interrogate	25/10/2022	25/10/2022
							F1901	Additional Surgery	A071102	Failure to Interrogate	25/10/2022	25/10/2022
					CT2407	Active implantable	E2403	No Clinical Signs, Symptoms or Conditions	A071102	Failure to Interrogate	25/10/2022	25/10/2022
							F1901	Additional Surgery	A071102	Failure to Interrogate	25/10/2022	25/10/2022
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A071102	Failure to Interrogate	25/10/2022	25/10/2022
							F1901	Additional Surgery	A071102	Failure to Interrogate	25/10/2022	25/10/2022
					CT2727	Active implant devices/ancillaries	E2403	No Clinical Signs, Symptoms or Conditions	A071102	Failure to Interrogate	25/10/2022	25/10/2022
							F1901	Additional Surgery	A071102	Failure to Interrogate	25/10/2022	25/10/2022
					CT2750	Implantable cardiac electrical stimulators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A071102	Failure to Interrogate	25/10/2022	25/10/2022
							F1901	Additional Surgery	A071102	Failure to Interrogate	25/10/2022	25/10/2022
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A071102	Failure to Interrogate	25/10/2022	25/10/2022
							F1901	Additional Surgery	A071102	Failure to Interrogate	25/10/2022	25/10/2022
					CT2985	Active	E2403	No Clinical Signs, Symptoms or Conditions	A071102	Failure to Interrogate	25/10/2022	25/10/2022
							F1901	Additional Surgery	A071102	Failure to Interrogate	25/10/2022	25/10/2022
					CT3001	Implantable metal	E2403	No Clinical Signs, Symptoms or Conditions	A071102	Failure to Interrogate	25/10/2022	25/10/2022
							F1901	Additional Surgery	A071102	Failure to Interrogate	25/10/2022	25/10/2022

					CT3002	Implantable synthetic polymer	E2403	No Clinical Signs, Symptoms or Conditions	A071102	Failure to Interrogate	25/10/2022	25/10/2022
							F1901	Additional Surgery	A071102	Failure to Interrogate	25/10/2022	25/10/2022
					CT3006	Implantable inorganic material	E2403	No Clinical Signs, Symptoms or Conditions	A071102	Failure to Interrogate	25/10/2022	25/10/2022
							F1901	Additional Surgery	A071102	Failure to Interrogate	25/10/2022	25/10/2022
					CT314	Electrically powered	E2403	No Clinical Signs, Symptoms or Conditions	A071102	Failure to Interrogate	25/10/2022	25/10/2022
							F1901	Additional Surgery	A071102	Failure to Interrogate	25/10/2022	25/10/2022
					CT321	Long-term surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A071102	Failure to Interrogate	25/10/2022	25/10/2022
							F1901	Additional Surgery	A071102	Failure to Interrogate	25/10/2022	25/10/2022
					CT449	Generators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A071102	Failure to Interrogate	25/10/2022	25/10/2022
							F1901	Additional Surgery	A071102	Failure to Interrogate	25/10/2022	25/10/2022
					CT981	Single use	E2403	No Clinical Signs, Symptoms or Conditions	A071102	Failure to Interrogate	25/10/2022	25/10/2022
							F1901	Additional Surgery	A071102	Failure to Interrogate	25/10/2022	25/10/2022
					CT983	Surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A071102	Failure to Interrogate	25/10/2022	25/10/2022
							F1901	Additional Surgery	A071102	Failure to Interrogate	25/10/2022	25/10/2022
11462	2022		47265	Dual-chamber implantable pacemaker, rate-responsive	CT141	Electrodes/Leads and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	08/10/2022	08/10/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	08/10/2022	08/10/2022
					CT145	Cardiovascular devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	08/10/2022	08/10/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	08/10/2022	08/10/2022
					CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	08/10/2022	08/10/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	08/10/2022	08/10/2022
					CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	08/10/2022	08/10/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	08/10/2022	08/10/2022
					CT233	Surgical	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	08/10/2022	08/10/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	08/10/2022	08/10/2022
					CT2406	Implantable	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	08/10/2022	08/10/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	08/10/2022	08/10/2022
					CT2407	Active implantable	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	08/10/2022	08/10/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	08/10/2022	08/10/2022
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	08/10/2022	08/10/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	08/10/2022	08/10/2022
					CT2727	Active implant devices/ancillaries	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	08/10/2022	08/10/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	08/10/2022	08/10/2022
					CT2750	Implantable cardiac electrical stimulators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	08/10/2022	08/10/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	08/10/2022	08/10/2022
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	08/10/2022	08/10/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	08/10/2022	08/10/2022
					CT2985	Active	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	08/10/2022	08/10/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	08/10/2022	08/10/2022
					CT3001	Implantable metal	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	08/10/2022	08/10/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	08/10/2022	08/10/2022
					CT3002	Implantable synthetic polymer	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	08/10/2022	08/10/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	08/10/2022	08/10/2022
					CT3006	Implantable inorganic material	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	08/10/2022	08/10/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	08/10/2022	08/10/2022
					CT314	Electrically powered	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	08/10/2022	08/10/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	08/10/2022	08/10/2022
					CT321	Long-term surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	08/10/2022	08/10/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	08/10/2022	08/10/2022
					CT449	Generators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	08/10/2022	08/10/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	08/10/2022	08/10/2022
					CT981	Single use	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	08/10/2022	08/10/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	08/10/2022	08/10/2022
					CT983	Surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	08/10/2022	08/10/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	08/10/2022	08/10/2022
11463	2022		47265	Dual-chamber implantable pacemaker, rate-responsive	CT141	Electrodes/Leads and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	08/10/2022	08/10/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	08/10/2022	08/10/2022
					CT145	Cardiovascular devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	08/10/2022	08/10/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	08/10/2022	08/10/2022
					CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	08/10/2022	08/10/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	08/10/2022	08/10/2022
					CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	08/10/2022	08/10/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	08/10/2022	08/10/2022
					CT233	Surgical	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	08/10/2022	08/10/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	08/10/2022	08/10/2022
					CT2406	Implantable	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	08/10/2022	08/10/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	08/10/2022	08/10/2022
					CT2407	Active implantable	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	08/10/2022	08/10/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	08/10/2022	08/10/2022
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	08/10/2022	08/10/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	08/10/2022	08/10/2022
					CT2727	Active implant devices/ancillaries	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	08/10/2022	08/10/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	08/10/2022	08/10/2022
					CT2750	Implantable cardiac electrical stimulators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	08/10/2022	08/10/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	08/10/2022	08/10/2022
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	08/10/2022	08/10/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	08/10/2022	08/10/2022
					CT2985	Active	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	08/10/2022	08/10/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	08/10/2022	08/10/2022
					CT3001	Implantable metal	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	08/10/2022	08/10/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	08/10/2022	08/10/2022
					CT3002	Implantable synthetic polymer	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	08/10/2022	08/10/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	08/10/2022	08/10/2022
					CT3006	Implantable inorganic material	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	08/10/2022	08/10/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	08/10/2022	08/10/2022
					CT314	Electrically powered	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	08/10/2022	08/10/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	08/10/2022	08/10/2022
					CT321	Long-term surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	08/10/2022	08/10/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	08/10/2022	08/10/2022

					CT449	Generators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	08/10/2022	08/10/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	08/10/2022	08/10/2022
					CT981	Single use	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	08/10/2022	08/10/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	08/10/2022	08/10/2022
					CT983	Surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	08/10/2022	08/10/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	08/10/2022	08/10/2022
11464	2022		47265	Dual-chamber implantable pacemaker, rate-responsive	CT141	Electrodes/Leads and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	08/10/2022	08/10/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	08/10/2022	08/10/2022
					CT145	Cardiovascular devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	08/10/2022	08/10/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	08/10/2022	08/10/2022
					CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	08/10/2022	08/10/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	08/10/2022	08/10/2022
					CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	08/10/2022	08/10/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	08/10/2022	08/10/2022
					CT233	Surgical	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	08/10/2022	08/10/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	08/10/2022	08/10/2022
					CT2406	Implantable	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	08/10/2022	08/10/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	08/10/2022	08/10/2022
					CT2407	Active implantable	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	08/10/2022	08/10/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	08/10/2022	08/10/2022
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	08/10/2022	08/10/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	08/10/2022	08/10/2022
					CT2727	Active implant devices/ancillaries	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	08/10/2022	08/10/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	08/10/2022	08/10/2022
					CT2750	Implantable cardiac electrical stimulators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	08/10/2022	08/10/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	08/10/2022	08/10/2022
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	08/10/2022	08/10/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	08/10/2022	08/10/2022
					CT2985	Active	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	08/10/2022	08/10/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	08/10/2022	08/10/2022
					CT3001	Implantable metal	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	08/10/2022	08/10/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	08/10/2022	08/10/2022
					CT3002	Implantable synthetic polymer	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	08/10/2022	08/10/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	08/10/2022	08/10/2022
					CT3006	Implantable inorganic material	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	08/10/2022	08/10/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	08/10/2022	08/10/2022
					CT314	Electrically powered	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	08/10/2022	08/10/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	08/10/2022	08/10/2022
					CT321	Long-term surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	08/10/2022	08/10/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	08/10/2022	08/10/2022
					CT449	Generators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	08/10/2022	08/10/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	08/10/2022	08/10/2022
					CT981	Single use	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	08/10/2022	08/10/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	08/10/2022	08/10/2022
					CT983	Surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	08/10/2022	08/10/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	08/10/2022	08/10/2022
11465	2022		47265	Dual-chamber implantable pacemaker, rate-responsive	CT141	Electrodes/Leads and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	08/10/2022	08/10/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	08/10/2022	08/10/2022
					CT145	Cardiovascular devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	08/10/2022	08/10/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	08/10/2022	08/10/2022
					CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	08/10/2022	08/10/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	08/10/2022	08/10/2022
					CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	08/10/2022	08/10/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	08/10/2022	08/10/2022
					CT233	Surgical	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	08/10/2022	08/10/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	08/10/2022	08/10/2022
					CT2406	Implantable	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	08/10/2022	08/10/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	08/10/2022	08/10/2022
					CT2407	Active implantable	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	08/10/2022	08/10/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	08/10/2022	08/10/2022
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	08/10/2022	08/10/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	08/10/2022	08/10/2022
					CT2727	Active implant devices/ancillaries	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	08/10/2022	08/10/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	08/10/2022	08/10/2022
					CT2750	Implantable cardiac electrical stimulators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	08/10/2022	08/10/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	08/10/2022	08/10/2022
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	08/10/2022	08/10/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	08/10/2022	08/10/2022
					CT2985	Active	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	08/10/2022	08/10/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	08/10/2022	08/10/2022
					CT3001	Implantable metal	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	08/10/2022	08/10/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	08/10/2022	08/10/2022
					CT3002	Implantable synthetic polymer	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	08/10/2022	08/10/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	08/10/2022	08/10/2022
					CT3006	Implantable inorganic material	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	08/10/2022	08/10/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	08/10/2022	08/10/2022
					CT314	Electrically powered	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	08/10/2022	08/10/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	08/10/2022	08/10/2022
					CT321	Long-term surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	08/10/2022	08/10/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	08/10/2022	08/10/2022
					CT449	Generators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	08/10/2022	08/10/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	08/10/2022	08/10/2022
					CT981	Single use	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	08/10/2022	08/10/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	08/10/2022	08/10/2022
					CT983	Surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	08/10/2022	08/10/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	08/10/2022	08/10/2022
11466	2022		47265	Dual-chamber implantable pacemaker, rate-responsive	CT141	Electrodes/Leads and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	26/10/2022	26/10/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	26/10/2022	26/10/2022

					CT145	Cardiovascular devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	26/10/2022	26/10/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	26/10/2022	26/10/2022
					CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	26/10/2022	26/10/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	26/10/2022	26/10/2022
					CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	26/10/2022	26/10/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	26/10/2022	26/10/2022
					CT233	Surgical	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	26/10/2022	26/10/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	26/10/2022	26/10/2022
					CT2406	Implantable	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	26/10/2022	26/10/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	26/10/2022	26/10/2022
					CT2407	Active implantable	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	26/10/2022	26/10/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	26/10/2022	26/10/2022
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	26/10/2022	26/10/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	26/10/2022	26/10/2022
					CT2727	Active implant devices/ancillaries	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	26/10/2022	26/10/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	26/10/2022	26/10/2022
					CT2750	Implantable cardiac electrical stimulators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	26/10/2022	26/10/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	26/10/2022	26/10/2022
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	26/10/2022	26/10/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	26/10/2022	26/10/2022
					CT2985	Active	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	26/10/2022	26/10/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	26/10/2022	26/10/2022
					CT3001	Implantable metal	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	26/10/2022	26/10/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	26/10/2022	26/10/2022
					CT3002	Implantable synthetic polymer	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	26/10/2022	26/10/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	26/10/2022	26/10/2022
					CT3006	Implantable inorganic material	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	26/10/2022	26/10/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	26/10/2022	26/10/2022
					CT314	Electrically powered	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	26/10/2022	26/10/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	26/10/2022	26/10/2022
					CT321	Long-term surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	26/10/2022	26/10/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	26/10/2022	26/10/2022
					CT449	Generators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	26/10/2022	26/10/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	26/10/2022	26/10/2022
					CT981	Single use	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	26/10/2022	26/10/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	26/10/2022	26/10/2022
					CT983	Surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	26/10/2022	26/10/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	26/10/2022	26/10/2022
11467	2022	All other reportable incidents	47265	Dual-chamber implantable pacemaker, rate-responsive	CT141	Electrodes/Leads and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	20/10/2022	20/10/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	20/10/2022	20/10/2022
					CT145	Cardiovascular devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	20/10/2022	20/10/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	20/10/2022	20/10/2022
					CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	20/10/2022	20/10/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	20/10/2022	20/10/2022
					CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	20/10/2022	20/10/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	20/10/2022	20/10/2022
					CT233	Surgical	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	20/10/2022	20/10/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	20/10/2022	20/10/2022
					CT2406	Implantable	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	20/10/2022	20/10/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	20/10/2022	20/10/2022
					CT2407	Active implantable	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	20/10/2022	20/10/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	20/10/2022	20/10/2022
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	20/10/2022	20/10/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	20/10/2022	20/10/2022
					CT2727	Active implant devices/ancillaries	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	20/10/2022	20/10/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	20/10/2022	20/10/2022
					CT2750	Implantable cardiac electrical stimulators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	20/10/2022	20/10/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	20/10/2022	20/10/2022
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	20/10/2022	20/10/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	20/10/2022	20/10/2022
					CT2985	Active	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	20/10/2022	20/10/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	20/10/2022	20/10/2022
					CT3001	Implantable metal	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	20/10/2022	20/10/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	20/10/2022	20/10/2022
					CT3002	Implantable synthetic polymer	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	20/10/2022	20/10/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	20/10/2022	20/10/2022
					CT3006	Implantable inorganic material	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	20/10/2022	20/10/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	20/10/2022	20/10/2022
					CT314	Electrically powered	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	20/10/2022	20/10/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	20/10/2022	20/10/2022
					CT321	Long-term surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	20/10/2022	20/10/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	20/10/2022	20/10/2022
					CT449	Generators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	20/10/2022	20/10/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	20/10/2022	20/10/2022
					CT981	Single use	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	20/10/2022	20/10/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	20/10/2022	20/10/2022
					CT983	Surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	20/10/2022	20/10/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	20/10/2022	20/10/2022
11468	2022	All other reportable incidents	35852	Single-chamber implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem	13/01/2023	13/01/2023
									A070504	Premature Discharge of Battery	13/01/2023	13/01/2023
									A0908	Incorrect, Inadequate or Imprecise Result or Readings	13/01/2023	13/01/2023
							F08	Hospitalization or Prolonged Hospitalization	A0705	Battery Problem	13/01/2023	13/01/2023
									A070504	Premature Discharge of Battery	13/01/2023	13/01/2023
									A0908	Incorrect, Inadequate or Imprecise Result or Readings	13/01/2023	13/01/2023
							F1905	Device Revision or Replacement	A0705	Battery Problem	13/01/2023	13/01/2023
									A070504	Premature Discharge of Battery	13/01/2023	13/01/2023

									A0908	Incorrect, Inadequate or Imprecise Result or Readings	13/01/2023	13/01/2023
11469	2022	All other reportable incidents	17882	Physiologic-monitoring defibrillation system	CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A0801	Failure to Calibrate		
							F27	Problem Identified During Non-Clinical Procedure	A0801	Failure to Calibrate		
11470	2022	All other reportable incidents	47267	Single-chamber implantable pacemaker, rate-responsive	CT141	Electrodes/Leads and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	31/10/2022	31/10/2022
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	31/10/2022	31/10/2022
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	31/10/2022	31/10/2022
					CT145	Cardiovascular devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	31/10/2022	31/10/2022
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	31/10/2022	31/10/2022
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	31/10/2022	31/10/2022
					CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	31/10/2022	31/10/2022
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	31/10/2022	31/10/2022
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	31/10/2022	31/10/2022
					CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	31/10/2022	31/10/2022
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	31/10/2022	31/10/2022
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	31/10/2022	31/10/2022
					CT233	Surgical	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	31/10/2022	31/10/2022
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	31/10/2022	31/10/2022
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	31/10/2022	31/10/2022
					CT2406	Implantable	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	31/10/2022	31/10/2022
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	31/10/2022	31/10/2022
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	31/10/2022	31/10/2022
					CT2407	Active implantable	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	31/10/2022	31/10/2022
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	31/10/2022	31/10/2022
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	31/10/2022	31/10/2022
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	31/10/2022	31/10/2022
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	31/10/2022	31/10/2022
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	31/10/2022	31/10/2022
					CT2727	Active implant devices/ancillaries	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	31/10/2022	31/10/2022
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	31/10/2022	31/10/2022
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	31/10/2022	31/10/2022
					CT2750	Implantable cardiac electrical stimulators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	31/10/2022	31/10/2022
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	31/10/2022	31/10/2022
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	31/10/2022	31/10/2022
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	31/10/2022	31/10/2022
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	31/10/2022	31/10/2022
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	31/10/2022	31/10/2022
					CT2985	Active	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	31/10/2022	31/10/2022
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	31/10/2022	31/10/2022
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	31/10/2022	31/10/2022
					CT3001	Implantable metal	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	31/10/2022	31/10/2022
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	31/10/2022	31/10/2022
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	31/10/2022	31/10/2022
					CT3002	Implantable synthetic polymer	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	31/10/2022	31/10/2022
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	31/10/2022	31/10/2022
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	31/10/2022	31/10/2022
					CT3006	Implantable inorganic material	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	31/10/2022	31/10/2022
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	31/10/2022	31/10/2022
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	31/10/2022	31/10/2022
					CT314	Electrically powered	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	31/10/2022	31/10/2022
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	31/10/2022	31/10/2022
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	31/10/2022	31/10/2022
					CT321	Long-term surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	31/10/2022	31/10/2022
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	31/10/2022	31/10/2022
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	31/10/2022	31/10/2022
					CT449	Generators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	31/10/2022	31/10/2022
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	31/10/2022	31/10/2022
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	31/10/2022	31/10/2022
					CT981	Single use	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	31/10/2022	31/10/2022
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	31/10/2022	31/10/2022
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	31/10/2022	31/10/2022
					CT983	Surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	31/10/2022	31/10/2022
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	31/10/2022	31/10/2022
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	31/10/2022	31/10/2022
11471	2022	All other reportable incidents			CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem		
							F26	No Health Consequences or Impact	A0705	Battery Problem		
11472	2022	All other reportable incidents	35852	Single-chamber implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
11473	2022	All other reportable incidents	37265	Dual-chamber implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery		
11474	2022	Unanticipated serious deterioration in state of health	17882	Physiologic-monitoring defibrillation system	CT1684	Cardiac pacemakers and associated devices	F26	No Health Consequences or Impact	A090811	Unable to Obtain Readings		
11475	2022		47265	Dual-chamber implantable pacemaker, rate-responsive	CT141	Electrodes/Leads and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	26/10/2022	26/10/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	26/10/2022	26/10/2022
					CT145	Cardiovascular devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	26/10/2022	26/10/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	26/10/2022	26/10/2022
					CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	26/10/2022	26/10/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	26/10/2022	26/10/2022
					CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	26/10/2022	26/10/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	26/10/2022	26/10/2022
					CT233	Surgical	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	26/10/2022	26/10/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	26/10/2022	26/10/2022
					CT2406	Implantable	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	26/10/2022	26/10/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	26/10/2022	26/10/2022
					CT2407	Active implantable	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	26/10/2022	26/10/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	26/10/2022	26/10/2022

					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	26/10/2022	26/10/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	26/10/2022	26/10/2022
					CT2727	Active implant devices/ancillaries	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	26/10/2022	26/10/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	26/10/2022	26/10/2022
					CT2750	Implantable cardiac electrical stimulators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	26/10/2022	26/10/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	26/10/2022	26/10/2022
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	26/10/2022	26/10/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	26/10/2022	26/10/2022
					CT2985	Active	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	26/10/2022	26/10/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	26/10/2022	26/10/2022
					CT3001	Implantable metal	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	26/10/2022	26/10/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	26/10/2022	26/10/2022
					CT3002	Implantable synthetic polymer	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	26/10/2022	26/10/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	26/10/2022	26/10/2022
					CT3006	Implantable inorganic material	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	26/10/2022	26/10/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	26/10/2022	26/10/2022
					CT314	Electrically powered	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	26/10/2022	26/10/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	26/10/2022	26/10/2022
					CT321	Long-term surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	26/10/2022	26/10/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	26/10/2022	26/10/2022
					CT449	Generators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	26/10/2022	26/10/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	26/10/2022	26/10/2022
					CT981	Single use	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	26/10/2022	26/10/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	26/10/2022	26/10/2022
					CT983	Surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	26/10/2022	26/10/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	26/10/2022	26/10/2022
11476	2022	All other reportable incidents	47265	Dual-chamber implantable pacemaker, rate-responsive	CT141	Electrodes/Leads and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070909	Oversensing		
									A090801	Signal Artifact/Noise		
							F26	No Health Consequences or Impact	A070909	Oversensing		
									A090801	Signal Artifact/Noise		
					CT145	Cardiovascular devices	E2403	No Clinical Signs, Symptoms or Conditions	A070909	Oversensing		
									A090801	Signal Artifact/Noise		
							F26	No Health Consequences or Impact	A070909	Oversensing		
									A090801	Signal Artifact/Noise		
					CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070909	Oversensing		
									A090801	Signal Artifact/Noise		
							F26	No Health Consequences or Impact	A070909	Oversensing		
									A090801	Signal Artifact/Noise		
					CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A070909	Oversensing		
									A090801	Signal Artifact/Noise		
							F26	No Health Consequences or Impact	A070909	Oversensing		
									A090801	Signal Artifact/Noise		
					CT233	Surgical	E2403	No Clinical Signs, Symptoms or Conditions	A070909	Oversensing		
									A090801	Signal Artifact/Noise		
							F26	No Health Consequences or Impact	A070909	Oversensing		
									A090801	Signal Artifact/Noise		
					CT2406	Implantable	E2403	No Clinical Signs, Symptoms or Conditions	A070909	Oversensing		
									A090801	Signal Artifact/Noise		
							F26	No Health Consequences or Impact	A070909	Oversensing		
									A090801	Signal Artifact/Noise		
					CT2407	Active implantable	E2403	No Clinical Signs, Symptoms or Conditions	A070909	Oversensing		
									A090801	Signal Artifact/Noise		
							F26	No Health Consequences or Impact	A070909	Oversensing		
									A090801	Signal Artifact/Noise		
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A070909	Oversensing		
									A090801	Signal Artifact/Noise		
							F26	No Health Consequences or Impact	A070909	Oversensing		
									A090801	Signal Artifact/Noise		
					CT2727	Active implant devices/ancillaries	E2403	No Clinical Signs, Symptoms or Conditions	A070909	Oversensing		
									A090801	Signal Artifact/Noise		
							F26	No Health Consequences or Impact	A070909	Oversensing		
									A090801	Signal Artifact/Noise		
					CT2750	Implantable cardiac electrical stimulators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070909	Oversensing		
									A090801	Signal Artifact/Noise		
							F26	No Health Consequences or Impact	A070909	Oversensing		
									A090801	Signal Artifact/Noise		
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A070909	Oversensing		
									A090801	Signal Artifact/Noise		
							F26	No Health Consequences or Impact	A070909	Oversensing		
									A090801	Signal Artifact/Noise		
					CT2985	Active	E2403	No Clinical Signs, Symptoms or Conditions	A070909	Oversensing		
									A090801	Signal Artifact/Noise		
							F26	No Health Consequences or Impact	A070909	Oversensing		
									A090801	Signal Artifact/Noise		
					CT3001	Implantable metal	E2403	No Clinical Signs, Symptoms or Conditions	A070909	Oversensing		
									A090801	Signal Artifact/Noise		
							F26	No Health Consequences or Impact	A070909	Oversensing		
									A090801	Signal Artifact/Noise		
					CT3002	Implantable synthetic polymer	E2403	No Clinical Signs, Symptoms or Conditions	A070909	Oversensing		
									A090801	Signal Artifact/Noise		
							F26	No Health Consequences or Impact	A070909	Oversensing		
									A090801	Signal Artifact/Noise		
					CT3006	Implantable inorganic material	E2403	No Clinical Signs, Symptoms or Conditions	A070909	Oversensing		
									A090801	Signal Artifact/Noise		
							F26	No Health Consequences or Impact	A070909	Oversensing		
									A090801	Signal Artifact/Noise		
					CT314	Electrically powered	E2403	No Clinical Signs, Symptoms or Conditions	A070909	Oversensing		
									A090801	Signal Artifact/Noise		

						F26	No Health Consequences or Impact	A070909	Oversensing			
								A090801	Signal Artifact/Noise			
					CT321	Long-term surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A070909	Oversensing		
									A090801	Signal Artifact/Noise		
							F26	No Health Consequences or Impact	A070909	Oversensing		
									A090801	Signal Artifact/Noise		
					CT449	Generators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070909	Oversensing		
									A090801	Signal Artifact/Noise		
							F26	No Health Consequences or Impact	A070909	Oversensing		
									A090801	Signal Artifact/Noise		
					CT981	Single use	E2403	No Clinical Signs, Symptoms or Conditions	A070909	Oversensing		
									A090801	Signal Artifact/Noise		
							F26	No Health Consequences or Impact	A070909	Oversensing		
									A090801	Signal Artifact/Noise		
					CT983	Surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A070909	Oversensing		
									A090801	Signal Artifact/Noise		
							F26	No Health Consequences or Impact	A070909	Oversensing		
									A090801	Signal Artifact/Noise		
11477	2022	All other reportable incidents	47267	Single-chamber implantable pacemaker, rate-responsive	CT141	Electrodes/Leads and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	06/10/2022	06/10/2022
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	06/10/2022	06/10/2022
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	06/10/2022	06/10/2022
					CT145	Cardiovascular devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	06/10/2022	06/10/2022
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	06/10/2022	06/10/2022
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	06/10/2022	06/10/2022
					CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	06/10/2022	06/10/2022
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	06/10/2022	06/10/2022
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	06/10/2022	06/10/2022
					CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	06/10/2022	06/10/2022
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	06/10/2022	06/10/2022
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	06/10/2022	06/10/2022
					CT233	Surgical	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	06/10/2022	06/10/2022
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	06/10/2022	06/10/2022
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	06/10/2022	06/10/2022
					CT2406	Implantable	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	06/10/2022	06/10/2022
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	06/10/2022	06/10/2022
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	06/10/2022	06/10/2022
					CT2407	Active implantable	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	06/10/2022	06/10/2022
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	06/10/2022	06/10/2022
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	06/10/2022	06/10/2022
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	06/10/2022	06/10/2022
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	06/10/2022	06/10/2022
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	06/10/2022	06/10/2022
					CT2727	Active implant devices/ancillaries	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	06/10/2022	06/10/2022
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	06/10/2022	06/10/2022
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	06/10/2022	06/10/2022
					CT2750	Implantable cardiac electrical stimulators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	06/10/2022	06/10/2022
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	06/10/2022	06/10/2022
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	06/10/2022	06/10/2022
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	06/10/2022	06/10/2022
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	06/10/2022	06/10/2022
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	06/10/2022	06/10/2022
					CT2985	Active	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	06/10/2022	06/10/2022
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	06/10/2022	06/10/2022
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	06/10/2022	06/10/2022
					CT3001	Implantable metal	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	06/10/2022	06/10/2022
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	06/10/2022	06/10/2022
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	06/10/2022	06/10/2022
					CT3002	Implantable synthetic polymer	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	06/10/2022	06/10/2022
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	06/10/2022	06/10/2022
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	06/10/2022	06/10/2022
					CT3006	Implantable inorganic material	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	06/10/2022	06/10/2022
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	06/10/2022	06/10/2022
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	06/10/2022	06/10/2022
					CT314	Electrically powered	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	06/10/2022	06/10/2022
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	06/10/2022	06/10/2022
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	06/10/2022	06/10/2022
					CT321	Long-term surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	06/10/2022	06/10/2022
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	06/10/2022	06/10/2022
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	06/10/2022	06/10/2022
					CT449	Generators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	06/10/2022	06/10/2022
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	06/10/2022	06/10/2022
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	06/10/2022	06/10/2022
					CT981	Single use	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	06/10/2022	06/10/2022
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	06/10/2022	06/10/2022
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	06/10/2022	06/10/2022
					CT983	Surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	06/10/2022	06/10/2022
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	06/10/2022	06/10/2022
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	06/10/2022	06/10/2022
11478	2022	All other reportable incidents	35852	Single-chamber implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	05/07/2023	05/07/2023
									A0908	Incorrect, Inadequate or Imprecise Result or Readings	05/07/2023	05/07/2023
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	05/07/2023	05/07/2023
									A0908	Incorrect, Inadequate or Imprecise Result or Readings	05/07/2023	05/07/2023
							F1903	Device Explantation	A070504	Premature Discharge of Battery	05/07/2023	05/07/2023
									A0908	Incorrect, Inadequate or Imprecise Result or Readings	05/07/2023	05/07/2023
11479	2022	All other reportable incidents	17882	Physiologic-monitoring defibrillation system	CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070803	Failure to Power Up		

							F27	Problem Identified During Non-Clinical Procedure	A070803	Failure to Power Up		
11480	2022	All other reportable incidents	47265	Dual-chamber implantable pacemaker, rate-responsive	CT141	Electrodes/Leads and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem		
									A070910	Undersensing		
							F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A0705	Battery Problem		
									A070910	Undersensing		
					CT145	Cardiovascular devices	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem		
									A070910	Undersensing		
							F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A0705	Battery Problem		
									A070910	Undersensing		
					CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem		
									A070910	Undersensing		
							F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A0705	Battery Problem		
									A070910	Undersensing		
					CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem		
									A070910	Undersensing		
							F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A0705	Battery Problem		
									A070910	Undersensing		
					CT233	Surgical	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem		
									A070910	Undersensing		
							F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A0705	Battery Problem		
									A070910	Undersensing		
					CT2406	Implantable	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem		
									A070910	Undersensing		
							F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A0705	Battery Problem		
									A070910	Undersensing		
					CT2407	Active implantable	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem		
									A070910	Undersensing		
							F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A0705	Battery Problem		
									A070910	Undersensing		
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem		
									A070910	Undersensing		
							F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A0705	Battery Problem		
									A070910	Undersensing		
					CT2727	Active implant devices/ancillaries	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem		
									A070910	Undersensing		
							F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A0705	Battery Problem		
									A070910	Undersensing		
					CT2750	Implantable cardiac electrical stimulators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem		
									A070910	Undersensing		
							F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A0705	Battery Problem		
									A070910	Undersensing		
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem		
									A070910	Undersensing		
							F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A0705	Battery Problem		
									A070910	Undersensing		
					CT2985	Active	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem		
									A070910	Undersensing		
							F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A0705	Battery Problem		
									A070910	Undersensing		
					CT3001	Implantable metal	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem		
									A070910	Undersensing		
							F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A0705	Battery Problem		
									A070910	Undersensing		
					CT3002	Implantable synthetic polymer	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem		
									A070910	Undersensing		
							F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A0705	Battery Problem		
									A070910	Undersensing		
					CT3006	Implantable inorganic material	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem		
									A070910	Undersensing		
							F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A0705	Battery Problem		
									A070910	Undersensing		
					CT314	Electrically powered	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem		
									A070910	Undersensing		
							F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A0705	Battery Problem		
									A070910	Undersensing		
					CT321	Long-term surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem		
									A070910	Undersensing		
							F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A0705	Battery Problem		
									A070910	Undersensing		
					CT449	Generators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem		
									A070910	Undersensing		
							F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A0705	Battery Problem		
									A070910	Undersensing		
					CT981	Single use	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem		
									A070910	Undersensing		
							F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A0705	Battery Problem		
									A070910	Undersensing		
					CT983	Surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem		
									A070910	Undersensing		
							F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A0705	Battery Problem		
									A070910	Undersensing		
11481	2022	All other reportable incidents	36241	Pacemaker/defibrillator lead	CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A072201	High Impedance		
							F26	No Health Consequences or Impact	A072201	High Impedance		
11482	2022	Unanticipated serious deterioration in state of health	60789	Intracardiac pacemaker	CT130	Catheters and associated devices						
					CT145	Cardiovascular devices						
					CT1684	Cardiac pacemakers and associated devices						
					CT1685	Cardiac pacemakers						
					CT233	Surgical						

					CT2406	Implantable							
					CT2407	Active implantable							
					CT257	Implantable cardiac pacemakers							
					CT2727	Active implant devices/ancillaries							
					CT2750	Implantable cardiac electrical stimulators and associated devices							
					CT2818	Catheterization/Cannulation/Fluid path devices							
					CT2866	Implantable cardiac electrical stimulators							
					CT2985	Active							
					CT3001	Implantable metal							
					CT3002	Implantable synthetic polymer							
					CT3006	Implantable inorganic material							
					CT314	Electrically powered							
					CT321	Long-term surgical invasive							
					CT981	Single use							
					CT983	Surgical invasive							
11483	2022	Unanticipated serious deterioration in state of health	60789	Intracardiac pacemaker	CT130	Catheters and associated devices			A0101	Patient-Device Incompatibility			
					CT145	Cardiovascular devices			A0101	Patient-Device Incompatibility			
					CT1684	Cardiac pacemakers and associated devices			A0101	Patient-Device Incompatibility			
					CT1685	Cardiac pacemakers			A0101	Patient-Device Incompatibility			
					CT233	Surgical			A0101	Patient-Device Incompatibility			
					CT2406	Implantable			A0101	Patient-Device Incompatibility			
					CT2407	Active implantable			A0101	Patient-Device Incompatibility			
					CT257	Implantable cardiac pacemakers			A0101	Patient-Device Incompatibility			
					CT2727	Active implant devices/ancillaries			A0101	Patient-Device Incompatibility			
					CT2750	Implantable cardiac electrical stimulators and associated devices			A0101	Patient-Device Incompatibility			
					CT2818	Catheterization/Cannulation/Fluid path devices			A0101	Patient-Device Incompatibility			
					CT2866	Implantable cardiac electrical stimulators			A0101	Patient-Device Incompatibility			
					CT2985	Active			A0101	Patient-Device Incompatibility			
					CT3001	Implantable metal			A0101	Patient-Device Incompatibility			
					CT3002	Implantable synthetic polymer			A0101	Patient-Device Incompatibility			
					CT3006	Implantable inorganic material			A0101	Patient-Device Incompatibility			
					CT314	Electrically powered			A0101	Patient-Device Incompatibility			
					CT321	Long-term surgical invasive			A0101	Patient-Device Incompatibility			
					CT981	Single use			A0101	Patient-Device Incompatibility			
					CT983	Surgical invasive			A0101	Patient-Device Incompatibility			
11484	2022	All other reportable incidents	47263	Cardiac resynchronization therapy implantable pacemaker	CT1684	Cardiac pacemakers and associated devices	E011903	Syncope/Fainting	A070909	Oversensing	13/10/2022	13/10/2022	
									A0712	Pacing Problem	13/10/2022	13/10/2022	
									A0908	Incorrect, Inadequate or Imprecise Result or Readings	13/10/2022	13/10/2022	
							E060101	Asystole	A070909	Oversensing	13/10/2022	13/10/2022	
									A0712	Pacing Problem	13/10/2022	13/10/2022	
									A0908	Incorrect, Inadequate or Imprecise Result or Readings	13/10/2022	13/10/2022	
							F08	Hospitalization or Prolonged Hospitalization	A070909	Oversensing	13/10/2022	13/10/2022	
									A0712	Pacing Problem	13/10/2022	13/10/2022	
									A0908	Incorrect, Inadequate or Imprecise Result or Readings	13/10/2022	13/10/2022	
							F1905	Device Revision or Replacement	A070909	Oversensing	13/10/2022	13/10/2022	
									A0712	Pacing Problem	13/10/2022	13/10/2022	
									A0908	Incorrect, Inadequate or Imprecise Result or Readings	13/10/2022	13/10/2022	
					CT1685	Cardiac pacemakers	E011903	Syncope/Fainting	A070909	Oversensing	13/10/2022	13/10/2022	
									A0712	Pacing Problem	13/10/2022	13/10/2022	
									A0908	Incorrect, Inadequate or Imprecise Result or Readings	13/10/2022	13/10/2022	
							E060101	Asystole	A070909	Oversensing	13/10/2022	13/10/2022	
									A0712	Pacing Problem	13/10/2022	13/10/2022	
									A0908	Incorrect, Inadequate or Imprecise Result or Readings	13/10/2022	13/10/2022	
							F08	Hospitalization or Prolonged Hospitalization	A070909	Oversensing	13/10/2022	13/10/2022	
									A0712	Pacing Problem	13/10/2022	13/10/2022	
									A0908	Incorrect, Inadequate or Imprecise Result or Readings	13/10/2022	13/10/2022	
							F1905	Device Revision or Replacement	A070909	Oversensing	13/10/2022	13/10/2022	
									A0712	Pacing Problem	13/10/2022	13/10/2022	
									A0908	Incorrect, Inadequate or Imprecise Result or Readings	13/10/2022	13/10/2022	
					CT257	Implantable cardiac pacemakers	E011903	Syncope/Fainting	A070909	Oversensing	13/10/2022	13/10/2022	
									A0712	Pacing Problem	13/10/2022	13/10/2022	
									A0908	Incorrect, Inadequate or Imprecise Result or Readings	13/10/2022	13/10/2022	
							E060101	Asystole	A070909	Oversensing	13/10/2022	13/10/2022	
									A0712	Pacing Problem	13/10/2022	13/10/2022	
									A0908	Incorrect, Inadequate or Imprecise Result or Readings	13/10/2022	13/10/2022	
							F08	Hospitalization or Prolonged Hospitalization	A070909	Oversensing	13/10/2022	13/10/2022	
									A0712	Pacing Problem	13/10/2022	13/10/2022	
									A0908	Incorrect, Inadequate or Imprecise Result or Readings	13/10/2022	13/10/2022	
							F1905	Device Revision or Replacement	A070909	Oversensing	13/10/2022	13/10/2022	
									A0712	Pacing Problem	13/10/2022	13/10/2022	
									A0908	Incorrect, Inadequate or Imprecise Result or Readings	13/10/2022	13/10/2022	
					CT2866	Implantable cardiac electrical stimulators	E011903	Syncope/Fainting	A070909	Oversensing	13/10/2022	13/10/2022	
									A0712	Pacing Problem	13/10/2022	13/10/2022	
									A0908	Incorrect, Inadequate or Imprecise Result or Readings	13/10/2022	13/10/2022	
							E060101	Asystole	A070909	Oversensing	13/10/2022	13/10/2022	
									A0712	Pacing Problem	13/10/2022	13/10/2022	

								A0908	Incorrect, Inadequate or Imprecise Result or Readings	13/10/2022	13/10/2022
						F08	Hospitalization or Prolonged Hospitalization	A070909	Oversensing	13/10/2022	13/10/2022
								A0712	Pacing Problem	13/10/2022	13/10/2022
								A0908	Incorrect, Inadequate or Imprecise Result or Readings	13/10/2022	13/10/2022
						F1905	Device Revision or Replacement	A070909	Oversensing	13/10/2022	13/10/2022
								A0712	Pacing Problem	13/10/2022	13/10/2022
								A0908	Incorrect, Inadequate or Imprecise Result or Readings	13/10/2022	13/10/2022
11485	2022	All other reportable incidents	47265	Dual-chamber implantable pacemaker, rate-responsive	CT141	Electrodes/Leads and associated devices	E2311	Discomfort	A0203	Defective Device	31/03/2022	31/03/2022
							F08	Hospitalization or Prolonged Hospitalization	A0203	Defective Device	31/03/2022	31/03/2022
							F1905	Device Revision or Replacement	A0203	Defective Device	31/03/2022	31/03/2022
					CT145	Cardiovascular devices	E2311	Discomfort	A0203	Defective Device	31/03/2022	31/03/2022
							F08	Hospitalization or Prolonged Hospitalization	A0203	Defective Device	31/03/2022	31/03/2022
							F1905	Device Revision or Replacement	A0203	Defective Device	31/03/2022	31/03/2022
					CT1684	Cardiac pacemakers and associated devices	E2311	Discomfort	A0203	Defective Device	31/03/2022	31/03/2022
							F08	Hospitalization or Prolonged Hospitalization	A0203	Defective Device	31/03/2022	31/03/2022
							F1905	Device Revision or Replacement	A0203	Defective Device	31/03/2022	31/03/2022
					CT1685	Cardiac pacemakers	E2311	Discomfort	A0203	Defective Device	31/03/2022	31/03/2022
							F08	Hospitalization or Prolonged Hospitalization	A0203	Defective Device	31/03/2022	31/03/2022
							F1905	Device Revision or Replacement	A0203	Defective Device	31/03/2022	31/03/2022
					CT233	Surgical	E2311	Discomfort	A0203	Defective Device	31/03/2022	31/03/2022
							F08	Hospitalization or Prolonged Hospitalization	A0203	Defective Device	31/03/2022	31/03/2022
							F1905	Device Revision or Replacement	A0203	Defective Device	31/03/2022	31/03/2022
					CT2406	Implantable	E2311	Discomfort	A0203	Defective Device	31/03/2022	31/03/2022
							F08	Hospitalization or Prolonged Hospitalization	A0203	Defective Device	31/03/2022	31/03/2022
							F1905	Device Revision or Replacement	A0203	Defective Device	31/03/2022	31/03/2022
					CT2407	Active implantable	E2311	Discomfort	A0203	Defective Device	31/03/2022	31/03/2022
							F08	Hospitalization or Prolonged Hospitalization	A0203	Defective Device	31/03/2022	31/03/2022
							F1905	Device Revision or Replacement	A0203	Defective Device	31/03/2022	31/03/2022
					CT257	Implantable cardiac pacemakers	E2311	Discomfort	A0203	Defective Device	31/03/2022	31/03/2022
							F08	Hospitalization or Prolonged Hospitalization	A0203	Defective Device	31/03/2022	31/03/2022
							F1905	Device Revision or Replacement	A0203	Defective Device	31/03/2022	31/03/2022
					CT2727	Active implant devices/ancillaries	E2311	Discomfort	A0203	Defective Device	31/03/2022	31/03/2022
							F08	Hospitalization or Prolonged Hospitalization	A0203	Defective Device	31/03/2022	31/03/2022
							F1905	Device Revision or Replacement	A0203	Defective Device	31/03/2022	31/03/2022
					CT2750	Implantable cardiac electrical stimulators and associated devices	E2311	Discomfort	A0203	Defective Device	31/03/2022	31/03/2022
							F08	Hospitalization or Prolonged Hospitalization	A0203	Defective Device	31/03/2022	31/03/2022
							F1905	Device Revision or Replacement	A0203	Defective Device	31/03/2022	31/03/2022
					CT2866	Implantable cardiac electrical stimulators	E2311	Discomfort	A0203	Defective Device	31/03/2022	31/03/2022
							F08	Hospitalization or Prolonged Hospitalization	A0203	Defective Device	31/03/2022	31/03/2022
							F1905	Device Revision or Replacement	A0203	Defective Device	31/03/2022	31/03/2022
					CT2985	Active	E2311	Discomfort	A0203	Defective Device	31/03/2022	31/03/2022
							F08	Hospitalization or Prolonged Hospitalization	A0203	Defective Device	31/03/2022	31/03/2022
							F1905	Device Revision or Replacement	A0203	Defective Device	31/03/2022	31/03/2022
					CT3001	Implantable metal	E2311	Discomfort	A0203	Defective Device	31/03/2022	31/03/2022
							F08	Hospitalization or Prolonged Hospitalization	A0203	Defective Device	31/03/2022	31/03/2022
							F1905	Device Revision or Replacement	A0203	Defective Device	31/03/2022	31/03/2022
					CT3002	Implantable synthetic polymer	E2311	Discomfort	A0203	Defective Device	31/03/2022	31/03/2022
							F08	Hospitalization or Prolonged Hospitalization	A0203	Defective Device	31/03/2022	31/03/2022
							F1905	Device Revision or Replacement	A0203	Defective Device	31/03/2022	31/03/2022
					CT3006	Implantable inorganic material	E2311	Discomfort	A0203	Defective Device	31/03/2022	31/03/2022
							F08	Hospitalization or Prolonged Hospitalization	A0203	Defective Device	31/03/2022	31/03/2022
							F1905	Device Revision or Replacement	A0203	Defective Device	31/03/2022	31/03/2022
					CT314	Electrically powered	E2311	Discomfort	A0203	Defective Device	31/03/2022	31/03/2022
							F08	Hospitalization or Prolonged Hospitalization	A0203	Defective Device	31/03/2022	31/03/2022
							F1905	Device Revision or Replacement	A0203	Defective Device	31/03/2022	31/03/2022
					CT321	Long-term surgical invasive	E2311	Discomfort	A0203	Defective Device	31/03/2022	31/03/2022
							F08	Hospitalization or Prolonged Hospitalization	A0203	Defective Device	31/03/2022	31/03/2022
							F1905	Device Revision or Replacement	A0203	Defective Device	31/03/2022	31/03/2022
					CT449	Generators and associated devices	E2311	Discomfort	A0203	Defective Device	31/03/2022	31/03/2022
							F08	Hospitalization or Prolonged Hospitalization	A0203	Defective Device	31/03/2022	31/03/2022
							F1905	Device Revision or Replacement	A0203	Defective Device	31/03/2022	31/03/2022
					CT981	Single use	E2311	Discomfort	A0203	Defective Device	31/03/2022	31/03/2022
							F08	Hospitalization or Prolonged Hospitalization	A0203	Defective Device	31/03/2022	31/03/2022
							F1905	Device Revision or Replacement	A0203	Defective Device	31/03/2022	31/03/2022
					CT983	Surgical invasive	E2311	Discomfort	A0203	Defective Device	31/03/2022	31/03/2022
							F08	Hospitalization or Prolonged Hospitalization	A0203	Defective Device	31/03/2022	31/03/2022
							F1905	Device Revision or Replacement	A0203	Defective Device	31/03/2022	31/03/2022
11486	2022	All other reportable incidents	35852	Single-chamber implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	29/11/2022	29/11/2022
									A0908	Incorrect, Inadequate or Imprecise Result or Readings	29/11/2022	29/11/2022
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	29/11/2022	29/11/2022
									A0908	Incorrect, Inadequate or Imprecise Result or Readings	29/11/2022	29/11/2022
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	29/11/2022	29/11/2022
									A0908	Incorrect, Inadequate or Imprecise Result or Readings	29/11/2022	29/11/2022
11487	2022		47265	Dual-chamber implantable pacemaker, rate-responsive	CT141	Electrodes/Leads and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	03/10/2022	03/10/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	03/10/2022	03/10/2022
					CT145	Cardiovascular devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	03/10/2022	03/10/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	03/10/2022	03/10/2022
					CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	03/10/2022	03/10/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	03/10/2022	03/10/2022
					CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	03/10/2022	03/10/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	03/10/2022	03/10/2022
					CT233	Surgical	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	03/10/2022	03/10/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	03/10/2022	03/10/2022

					CT2406	Implantable	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	03/10/2022	03/10/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	03/10/2022	03/10/2022
					CT2407	Active implantable	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	03/10/2022	03/10/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	03/10/2022	03/10/2022
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	03/10/2022	03/10/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	03/10/2022	03/10/2022
					CT2727	Active implant devices/ancillaries	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	03/10/2022	03/10/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	03/10/2022	03/10/2022
					CT2750	Implantable cardiac electrical stimulators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	03/10/2022	03/10/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	03/10/2022	03/10/2022
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	03/10/2022	03/10/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	03/10/2022	03/10/2022
					CT2985	Active	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	03/10/2022	03/10/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	03/10/2022	03/10/2022
					CT3001	Implantable metal	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	03/10/2022	03/10/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	03/10/2022	03/10/2022
					CT3002	Implantable synthetic polymer	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	03/10/2022	03/10/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	03/10/2022	03/10/2022
					CT3006	Implantable inorganic material	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	03/10/2022	03/10/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	03/10/2022	03/10/2022
					CT314	Electrically powered	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	03/10/2022	03/10/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	03/10/2022	03/10/2022
					CT321	Long-term surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	03/10/2022	03/10/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	03/10/2022	03/10/2022
					CT449	Generators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	03/10/2022	03/10/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	03/10/2022	03/10/2022
					CT981	Single use	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	03/10/2022	03/10/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	03/10/2022	03/10/2022
					CT983	Surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	03/10/2022	03/10/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	03/10/2022	03/10/2022
11488	2022		47265	Dual-chamber implantable pacemaker, rate-responsive	CT141	Electrodes/Leads and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	12/10/2022	12/10/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	12/10/2022	12/10/2022
					CT145	Cardiovascular devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	12/10/2022	12/10/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	12/10/2022	12/10/2022
					CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	12/10/2022	12/10/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	12/10/2022	12/10/2022
					CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	12/10/2022	12/10/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	12/10/2022	12/10/2022
					CT233	Surgical	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	12/10/2022	12/10/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	12/10/2022	12/10/2022
					CT2406	Implantable	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	12/10/2022	12/10/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	12/10/2022	12/10/2022
					CT2407	Active implantable	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	12/10/2022	12/10/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	12/10/2022	12/10/2022
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	12/10/2022	12/10/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	12/10/2022	12/10/2022
					CT2727	Active implant devices/ancillaries	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	12/10/2022	12/10/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	12/10/2022	12/10/2022
					CT2750	Implantable cardiac electrical stimulators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	12/10/2022	12/10/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	12/10/2022	12/10/2022
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	12/10/2022	12/10/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	12/10/2022	12/10/2022
					CT2985	Active	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	12/10/2022	12/10/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	12/10/2022	12/10/2022
					CT3001	Implantable metal	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	12/10/2022	12/10/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	12/10/2022	12/10/2022
					CT3002	Implantable synthetic polymer	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	12/10/2022	12/10/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	12/10/2022	12/10/2022
					CT3006	Implantable inorganic material	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	12/10/2022	12/10/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	12/10/2022	12/10/2022
					CT314	Electrically powered	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	12/10/2022	12/10/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	12/10/2022	12/10/2022
					CT321	Long-term surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	12/10/2022	12/10/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	12/10/2022	12/10/2022
					CT449	Generators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	12/10/2022	12/10/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	12/10/2022	12/10/2022
					CT981	Single use	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	12/10/2022	12/10/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	12/10/2022	12/10/2022
					CT983	Surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	12/10/2022	12/10/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	12/10/2022	12/10/2022
11489	2022		35852	Single-chamber implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A070912	Incorrect Interpretation of Signal		
									A071302	Inappropriate/Inadequate Shock/Stimulation		
							F26	No Health Consequences or Impact	A070912	Incorrect Interpretation of Signal		
									A071302	Inappropriate/Inadequate Shock/Stimulation		
11490	2022	All other reportable incidents	47265	Dual-chamber implantable pacemaker, rate-responsive	CT141	Electrodes/Leads and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	13/10/2022	13/10/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	13/10/2022	13/10/2022
					CT145	Cardiovascular devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	13/10/2022	13/10/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	13/10/2022	13/10/2022
					CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	13/10/2022	13/10/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	13/10/2022	13/10/2022
					CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	13/10/2022	13/10/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	13/10/2022	13/10/2022
					CT233	Surgical	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	13/10/2022	13/10/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	13/10/2022	13/10/2022
					CT2406	Implantable	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	13/10/2022	13/10/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	13/10/2022	13/10/2022
					CT2407	Active implantable	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	13/10/2022	13/10/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	13/10/2022	13/10/2022

					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	13/10/2022	13/10/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	13/10/2022	13/10/2022
					CT2727	Active implant devices/ancillaries	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	13/10/2022	13/10/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	13/10/2022	13/10/2022
					CT2750	Implantable cardiac electrical stimulators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	13/10/2022	13/10/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	13/10/2022	13/10/2022
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	13/10/2022	13/10/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	13/10/2022	13/10/2022
					CT2985	Active	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	13/10/2022	13/10/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	13/10/2022	13/10/2022
					CT3001	Implantable metal	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	13/10/2022	13/10/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	13/10/2022	13/10/2022
					CT3002	Implantable synthetic polymer	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	13/10/2022	13/10/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	13/10/2022	13/10/2022
					CT3006	Implantable inorganic material	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	13/10/2022	13/10/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	13/10/2022	13/10/2022
					CT314	Electrically powered	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	13/10/2022	13/10/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	13/10/2022	13/10/2022
					CT321	Long-term surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	13/10/2022	13/10/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	13/10/2022	13/10/2022
					CT449	Generators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	13/10/2022	13/10/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	13/10/2022	13/10/2022
					CT981	Single use	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	13/10/2022	13/10/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	13/10/2022	13/10/2022
					CT983	Surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	13/10/2022	13/10/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	13/10/2022	13/10/2022
11491	2022	All other reportable incidents	47265	Dual-chamber implantable pacemaker, rate-responsive	CT141	Electrodes/Leads and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	11/10/2022	11/10/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	11/10/2022	11/10/2022
					CT145	Cardiovascular devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	11/10/2022	11/10/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	11/10/2022	11/10/2022
					CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	11/10/2022	11/10/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	11/10/2022	11/10/2022
					CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	11/10/2022	11/10/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	11/10/2022	11/10/2022
					CT233	Surgical	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	11/10/2022	11/10/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	11/10/2022	11/10/2022
					CT2406	Implantable	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	11/10/2022	11/10/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	11/10/2022	11/10/2022
					CT2407	Active implantable	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	11/10/2022	11/10/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	11/10/2022	11/10/2022
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	11/10/2022	11/10/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	11/10/2022	11/10/2022
					CT2727	Active implant devices/ancillaries	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	11/10/2022	11/10/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	11/10/2022	11/10/2022
					CT2750	Implantable cardiac electrical stimulators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	11/10/2022	11/10/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	11/10/2022	11/10/2022
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	11/10/2022	11/10/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	11/10/2022	11/10/2022
					CT2985	Active	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	11/10/2022	11/10/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	11/10/2022	11/10/2022
					CT3001	Implantable metal	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	11/10/2022	11/10/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	11/10/2022	11/10/2022
					CT3002	Implantable synthetic polymer	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	11/10/2022	11/10/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	11/10/2022	11/10/2022
					CT3006	Implantable inorganic material	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	11/10/2022	11/10/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	11/10/2022	11/10/2022
					CT314	Electrically powered	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	11/10/2022	11/10/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	11/10/2022	11/10/2022
					CT321	Long-term surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	11/10/2022	11/10/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	11/10/2022	11/10/2022
					CT449	Generators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	11/10/2022	11/10/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	11/10/2022	11/10/2022
					CT981	Single use	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	11/10/2022	11/10/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	11/10/2022	11/10/2022
					CT983	Surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	11/10/2022	11/10/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	11/10/2022	11/10/2022
11492	2022		47265	Dual-chamber implantable pacemaker, rate-responsive	CT141	Electrodes/Leads and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	29/09/2022	29/09/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	29/09/2022	29/09/2022
					CT145	Cardiovascular devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	29/09/2022	29/09/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	29/09/2022	29/09/2022
					CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	29/09/2022	29/09/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	29/09/2022	29/09/2022
					CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	29/09/2022	29/09/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	29/09/2022	29/09/2022
					CT233	Surgical	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	29/09/2022	29/09/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	29/09/2022	29/09/2022
					CT2406	Implantable	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	29/09/2022	29/09/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	29/09/2022	29/09/2022
					CT2407	Active implantable	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	29/09/2022	29/09/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	29/09/2022	29/09/2022
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	29/09/2022	29/09/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	29/09/2022	29/09/2022
					CT2727	Active implant devices/ancillaries	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	29/09/2022	29/09/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	29/09/2022	29/09/2022
					CT2750	Implantable cardiac electrical stimulators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	29/09/2022	29/09/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	29/09/2022	29/09/2022
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	29/09/2022	29/09/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	29/09/2022	29/09/2022

					CT2985	Active	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	29/09/2022	29/09/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	29/09/2022	29/09/2022
					CT3001	Implantable metal	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	29/09/2022	29/09/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	29/09/2022	29/09/2022
					CT3002	Implantable synthetic polymer	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	29/09/2022	29/09/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	29/09/2022	29/09/2022
					CT3006	Implantable inorganic material	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	29/09/2022	29/09/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	29/09/2022	29/09/2022
					CT314	Electrically powered	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	29/09/2022	29/09/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	29/09/2022	29/09/2022
					CT321	Long-term surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	29/09/2022	29/09/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	29/09/2022	29/09/2022
					CT449	Generators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	29/09/2022	29/09/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	29/09/2022	29/09/2022
					CT981	Single use	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	29/09/2022	29/09/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	29/09/2022	29/09/2022
					CT983	Surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	29/09/2022	29/09/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	29/09/2022	29/09/2022
11493	2022		47265	Dual-chamber implantable pacemaker, rate-responsive	CT141	Electrodes/Leads and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	26/10/2022	26/10/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	26/10/2022	26/10/2022
					CT145	Cardiovascular devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	26/10/2022	26/10/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	26/10/2022	26/10/2022
					CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	26/10/2022	26/10/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	26/10/2022	26/10/2022
					CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	26/10/2022	26/10/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	26/10/2022	26/10/2022
					CT233	Surgical	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	26/10/2022	26/10/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	26/10/2022	26/10/2022
					CT2406	Implantable	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	26/10/2022	26/10/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	26/10/2022	26/10/2022
					CT2407	Active implantable	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	26/10/2022	26/10/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	26/10/2022	26/10/2022
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	26/10/2022	26/10/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	26/10/2022	26/10/2022
					CT2727	Active implant devices/ancillaries	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	26/10/2022	26/10/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	26/10/2022	26/10/2022
					CT2750	Implantable cardiac electrical stimulators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	26/10/2022	26/10/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	26/10/2022	26/10/2022
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	26/10/2022	26/10/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	26/10/2022	26/10/2022
					CT2985	Active	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	26/10/2022	26/10/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	26/10/2022	26/10/2022
					CT3001	Implantable metal	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	26/10/2022	26/10/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	26/10/2022	26/10/2022
					CT3002	Implantable synthetic polymer	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	26/10/2022	26/10/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	26/10/2022	26/10/2022
					CT3006	Implantable inorganic material	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	26/10/2022	26/10/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	26/10/2022	26/10/2022
					CT314	Electrically powered	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	26/10/2022	26/10/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	26/10/2022	26/10/2022
					CT321	Long-term surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	26/10/2022	26/10/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	26/10/2022	26/10/2022
					CT449	Generators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	26/10/2022	26/10/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	26/10/2022	26/10/2022
					CT981	Single use	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	26/10/2022	26/10/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	26/10/2022	26/10/2022
					CT983	Surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	26/10/2022	26/10/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	26/10/2022	26/10/2022
11494	2022	All other reportable incidents	35852	Single-chamber implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E020201	Anxiety	A070504	Premature Discharge of Battery	21/10/2022	21/10/2022
									A0908	Incorrect, Inadequate or Imprecise Result or Readings	21/10/2022	21/10/2022
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	21/10/2022	21/10/2022
									A0908	Incorrect, Inadequate or Imprecise Result or Readings	21/10/2022	21/10/2022
							F11	Non-Serious Injury/ Illness/ Impairment	A070504	Premature Discharge of Battery	21/10/2022	21/10/2022
									A0908	Incorrect, Inadequate or Imprecise Result or Readings	21/10/2022	21/10/2022
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	21/10/2022	21/10/2022
									A0908	Incorrect, Inadequate or Imprecise Result or Readings	21/10/2022	21/10/2022
11495	2022	Unanticipated serious deterioration in state of health	60789	Intracardiac pacemaker	CT130	Catheters and associated devices						
					CT145	Cardiovascular devices						
					CT1684	Cardiac pacemakers and associated devices						
					CT1685	Cardiac pacemakers						
					CT233	Surgical						
					CT2406	Implantable						
					CT2407	Active implantable						
					CT257	Implantable cardiac pacemakers						
					CT2727	Active implant devices/ancillaries						
					CT2750	Implantable cardiac electrical stimulators and associated devices						
					CT2818	Catheterization/Cannulation/Fluid path devices						
					CT2866	Implantable cardiac electrical stimulators						
					CT2985	Active						
					CT3001	Implantable metal						
					CT3002	Implantable synthetic polymer						
					CT3006	Implantable inorganic material						
					CT314	Electrically powered						
					CT321	Long-term surgical invasive						

					CT981	Single use							
					CT983	Surgical invasive							
11496	2022	Unanticipated serious deterioration in state of health	60789	Intracardiac pacemaker	CT130	Catheters and associated devices			A26	Insufficient Device Problem Information			
					CT145	Cardiovascular devices			A26	Insufficient Device Problem Information			
					CT1684	Cardiac pacemakers and associated devices			A26	Insufficient Device Problem Information			
					CT1685	Cardiac pacemakers			A26	Insufficient Device Problem Information			
					CT233	Surgical			A26	Insufficient Device Problem Information			
					CT2406	Implantable			A26	Insufficient Device Problem Information			
					CT2407	Active implantable			A26	Insufficient Device Problem Information			
					CT257	Implantable cardiac pacemakers			A26	Insufficient Device Problem Information			
					CT2727	Active implant devices/ancillaries			A26	Insufficient Device Problem Information			
					CT2750	Implantable cardiac electrical stimulators and associated devices			A26	Insufficient Device Problem Information			
					CT2818	Catheterization/Cannulation/Fluid path devices			A26	Insufficient Device Problem Information			
					CT2866	Implantable cardiac electrical stimulators			A26	Insufficient Device Problem Information			
					CT2985	Active			A26	Insufficient Device Problem Information			
					CT3001	Implantable metal			A26	Insufficient Device Problem Information			
					CT3002	Implantable synthetic polymer			A26	Insufficient Device Problem Information			
					CT3006	Implantable inorganic material			A26	Insufficient Device Problem Information			
					CT314	Electrically powered			A26	Insufficient Device Problem Information			
					CT321	Long-term surgical invasive			A26	Insufficient Device Problem Information			
					CT981	Single use			A26	Insufficient Device Problem Information			
					CT983	Surgical invasive			A26	Insufficient Device Problem Information			
11497	2022	All other reportable incidents	47265	Dual-chamber implantable pacemaker, rate-responsive	CT141	Electrodes/Leads and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem	11/11/2022	11/11/2022	
									A070504	Premature Discharge of Battery	11/11/2022	11/11/2022	
							F08	Hospitalization or Prolonged Hospitalization	A0705	Battery Problem	11/11/2022	11/11/2022	
									A070504	Premature Discharge of Battery	11/11/2022	11/11/2022	
							F1905	Device Revision or Replacement	A0705	Battery Problem	11/11/2022	11/11/2022	
									A070504	Premature Discharge of Battery	11/11/2022	11/11/2022	
					CT145	Cardiovascular devices	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem	11/11/2022	11/11/2022	
									A070504	Premature Discharge of Battery	11/11/2022	11/11/2022	
							F08	Hospitalization or Prolonged Hospitalization	A0705	Battery Problem	11/11/2022	11/11/2022	
									A070504	Premature Discharge of Battery	11/11/2022	11/11/2022	
							F1905	Device Revision or Replacement	A0705	Battery Problem	11/11/2022	11/11/2022	
									A070504	Premature Discharge of Battery	11/11/2022	11/11/2022	
					CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem	11/11/2022	11/11/2022	
									A070504	Premature Discharge of Battery	11/11/2022	11/11/2022	
							F08	Hospitalization or Prolonged Hospitalization	A0705	Battery Problem	11/11/2022	11/11/2022	
									A070504	Premature Discharge of Battery	11/11/2022	11/11/2022	
							F1905	Device Revision or Replacement	A0705	Battery Problem	11/11/2022	11/11/2022	
									A070504	Premature Discharge of Battery	11/11/2022	11/11/2022	
					CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem	11/11/2022	11/11/2022	
									A070504	Premature Discharge of Battery	11/11/2022	11/11/2022	
							F08	Hospitalization or Prolonged Hospitalization	A0705	Battery Problem	11/11/2022	11/11/2022	
									A070504	Premature Discharge of Battery	11/11/2022	11/11/2022	
							F1905	Device Revision or Replacement	A0705	Battery Problem	11/11/2022	11/11/2022	
									A070504	Premature Discharge of Battery	11/11/2022	11/11/2022	
					CT233	Surgical	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem	11/11/2022	11/11/2022	
									A070504	Premature Discharge of Battery	11/11/2022	11/11/2022	
							F08	Hospitalization or Prolonged Hospitalization	A0705	Battery Problem	11/11/2022	11/11/2022	
									A070504	Premature Discharge of Battery	11/11/2022	11/11/2022	
							F1905	Device Revision or Replacement	A0705	Battery Problem	11/11/2022	11/11/2022	
									A070504	Premature Discharge of Battery	11/11/2022	11/11/2022	
					CT2406	Implantable	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem	11/11/2022	11/11/2022	
									A070504	Premature Discharge of Battery	11/11/2022	11/11/2022	
							F08	Hospitalization or Prolonged Hospitalization	A0705	Battery Problem	11/11/2022	11/11/2022	
									A070504	Premature Discharge of Battery	11/11/2022	11/11/2022	
							F1905	Device Revision or Replacement	A0705	Battery Problem	11/11/2022	11/11/2022	
									A070504	Premature Discharge of Battery	11/11/2022	11/11/2022	
					CT2407	Active implantable	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem	11/11/2022	11/11/2022	
									A070504	Premature Discharge of Battery	11/11/2022	11/11/2022	
							F08	Hospitalization or Prolonged Hospitalization	A0705	Battery Problem	11/11/2022	11/11/2022	
									A070504	Premature Discharge of Battery	11/11/2022	11/11/2022	
							F1905	Device Revision or Replacement	A0705	Battery Problem	11/11/2022	11/11/2022	
									A070504	Premature Discharge of Battery	11/11/2022	11/11/2022	
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem	11/11/2022	11/11/2022	
									A070504	Premature Discharge of Battery	11/11/2022	11/11/2022	
							F08	Hospitalization or Prolonged Hospitalization	A0705	Battery Problem	11/11/2022	11/11/2022	
									A070504	Premature Discharge of Battery	11/11/2022	11/11/2022	
							F1905	Device Revision or Replacement	A0705	Battery Problem	11/11/2022	11/11/2022	
									A070504	Premature Discharge of Battery	11/11/2022	11/11/2022	
					CT2727	Active implant devices/ancillaries	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem	11/11/2022	11/11/2022	
									A070504	Premature Discharge of Battery	11/11/2022	11/11/2022	
							F08	Hospitalization or Prolonged Hospitalization	A0705	Battery Problem	11/11/2022	11/11/2022	
									A070504	Premature Discharge of Battery	11/11/2022	11/11/2022	
							F1905	Device Revision or Replacement	A0705	Battery Problem	11/11/2022	11/11/2022	
									A070504	Premature Discharge of Battery	11/11/2022	11/11/2022	
					CT2750	Implantable cardiac electrical stimulators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem	11/11/2022	11/11/2022	
									A070504	Premature Discharge of Battery	11/11/2022	11/11/2022	
							F08	Hospitalization or Prolonged Hospitalization	A0705	Battery Problem	11/11/2022	11/11/2022	
									A070504	Premature Discharge of Battery	11/11/2022	11/11/2022	
							F1905	Device Revision or Replacement	A0705	Battery Problem	11/11/2022	11/11/2022	
									A070504	Premature Discharge of Battery	11/11/2022	11/11/2022	
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem	11/11/2022	11/11/2022	
									A070504	Premature Discharge of Battery	11/11/2022	11/11/2022	
							F08	Hospitalization or Prolonged Hospitalization	A0705	Battery Problem	11/11/2022	11/11/2022	
									A070504	Premature Discharge of Battery	11/11/2022	11/11/2022	
							F1905	Device Revision or Replacement	A0705	Battery Problem	11/11/2022	11/11/2022	

									A070504	Premature Discharge of Battery	11/11/2022	11/11/2022
					CT2985	Active	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem	11/11/2022	11/11/2022
									A070504	Premature Discharge of Battery	11/11/2022	11/11/2022
							F08	Hospitalization or Prolonged Hospitalization	A0705	Battery Problem	11/11/2022	11/11/2022
									A070504	Premature Discharge of Battery	11/11/2022	11/11/2022
							F1905	Device Revision or Replacement	A0705	Battery Problem	11/11/2022	11/11/2022
									A070504	Premature Discharge of Battery	11/11/2022	11/11/2022
					CT3001	Implantable metal	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem	11/11/2022	11/11/2022
									A070504	Premature Discharge of Battery	11/11/2022	11/11/2022
							F08	Hospitalization or Prolonged Hospitalization	A0705	Battery Problem	11/11/2022	11/11/2022
									A070504	Premature Discharge of Battery	11/11/2022	11/11/2022
							F1905	Device Revision or Replacement	A0705	Battery Problem	11/11/2022	11/11/2022
									A070504	Premature Discharge of Battery	11/11/2022	11/11/2022
					CT3002	Implantable synthetic polymer	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem	11/11/2022	11/11/2022
									A070504	Premature Discharge of Battery	11/11/2022	11/11/2022
							F08	Hospitalization or Prolonged Hospitalization	A0705	Battery Problem	11/11/2022	11/11/2022
									A070504	Premature Discharge of Battery	11/11/2022	11/11/2022
							F1905	Device Revision or Replacement	A0705	Battery Problem	11/11/2022	11/11/2022
									A070504	Premature Discharge of Battery	11/11/2022	11/11/2022
					CT3006	Implantable inorganic material	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem	11/11/2022	11/11/2022
									A070504	Premature Discharge of Battery	11/11/2022	11/11/2022
							F08	Hospitalization or Prolonged Hospitalization	A0705	Battery Problem	11/11/2022	11/11/2022
									A070504	Premature Discharge of Battery	11/11/2022	11/11/2022
							F1905	Device Revision or Replacement	A0705	Battery Problem	11/11/2022	11/11/2022
									A070504	Premature Discharge of Battery	11/11/2022	11/11/2022
					CT314	Electrically powered	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem	11/11/2022	11/11/2022
									A070504	Premature Discharge of Battery	11/11/2022	11/11/2022
							F08	Hospitalization or Prolonged Hospitalization	A0705	Battery Problem	11/11/2022	11/11/2022
									A070504	Premature Discharge of Battery	11/11/2022	11/11/2022
							F1905	Device Revision or Replacement	A0705	Battery Problem	11/11/2022	11/11/2022
									A070504	Premature Discharge of Battery	11/11/2022	11/11/2022
					CT321	Long-term surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem	11/11/2022	11/11/2022
									A070504	Premature Discharge of Battery	11/11/2022	11/11/2022
							F08	Hospitalization or Prolonged Hospitalization	A0705	Battery Problem	11/11/2022	11/11/2022
									A070504	Premature Discharge of Battery	11/11/2022	11/11/2022
							F1905	Device Revision or Replacement	A0705	Battery Problem	11/11/2022	11/11/2022
									A070504	Premature Discharge of Battery	11/11/2022	11/11/2022
					CT449	Generators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem	11/11/2022	11/11/2022
									A070504	Premature Discharge of Battery	11/11/2022	11/11/2022
							F08	Hospitalization or Prolonged Hospitalization	A0705	Battery Problem	11/11/2022	11/11/2022
									A070504	Premature Discharge of Battery	11/11/2022	11/11/2022
							F1905	Device Revision or Replacement	A0705	Battery Problem	11/11/2022	11/11/2022
									A070504	Premature Discharge of Battery	11/11/2022	11/11/2022
					CT981	Single use	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem	11/11/2022	11/11/2022
									A070504	Premature Discharge of Battery	11/11/2022	11/11/2022
							F08	Hospitalization or Prolonged Hospitalization	A0705	Battery Problem	11/11/2022	11/11/2022
									A070504	Premature Discharge of Battery	11/11/2022	11/11/2022
							F1905	Device Revision or Replacement	A0705	Battery Problem	11/11/2022	11/11/2022
									A070504	Premature Discharge of Battery	11/11/2022	11/11/2022
					CT983	Surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem	11/11/2022	11/11/2022
									A070504	Premature Discharge of Battery	11/11/2022	11/11/2022
							F08	Hospitalization or Prolonged Hospitalization	A0705	Battery Problem	11/11/2022	11/11/2022
									A070504	Premature Discharge of Battery	11/11/2022	11/11/2022
							F1905	Device Revision or Replacement	A0705	Battery Problem	11/11/2022	11/11/2022
									A070504	Premature Discharge of Battery	11/11/2022	11/11/2022
11498	2022	All other reportable incidents	17882	Physiologic-monitoring defibrillation system	CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070803	Failure to Power Up		
							F27	Problem Identified During Non-Clinical Procedure	A070803	Failure to Power Up		
11499	2022	All other reportable incidents	17882	Physiologic-monitoring defibrillation system	CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A071301	Failure to Deliver Shock/Stimulation		
							F26	No Health Consequences or Impact	A071301	Failure to Deliver Shock/Stimulation		
11500	2022	All other reportable incidents	17882	Physiologic-monitoring defibrillation system	CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A071301	Failure to Deliver Shock/Stimulation		
							F27	Problem Identified During Non-Clinical Procedure	A071301	Failure to Deliver Shock/Stimulation		
11501	2022	All other reportable incidents	37265	Dual-chamber implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A0709	Device Sensing Problem		
									A070909	Oversensing		
									A0722	Impedance Problem		
									A090801	Signal Artifact/Noise		
							F26	No Health Consequences or Impact	A0709	Device Sensing Problem		
									A070909	Oversensing		
									A0722	Impedance Problem		
									A090801	Signal Artifact/Noise		
11502	2022	All other reportable incidents	36241	Pacemaker/defibrillator lead	CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A072201	High Impedance		
							F26	No Health Consequences or Impact	A072201	High Impedance		
11503	2022	All other reportable incidents	37265	Dual-chamber implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A072201	High Impedance		
							F26	No Health Consequences or Impact	A072201	High Impedance		
11504	2022	All other reportable incidents	37265	Dual-chamber implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A0722	Impedance Problem		
									A072201	High Impedance		
							F26	No Health Consequences or Impact	A0722	Impedance Problem		
									A072201	High Impedance		
11505	2022	All other reportable incidents	36241	Pacemaker/defibrillator lead	CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A072201	High Impedance		
							F26	No Health Consequences or Impact	A072201	High Impedance		
11506	2022	All other reportable incidents	35852	Single-chamber implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E2104	Electric Shock	A070909	Oversensing		
									A071302	Inappropriate/Inadequate Shock/Stimulation		
									A090801	Signal Artifact/Noise		
									A1301	Failure to Read Input Signal		
							F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A070909	Oversensing		
									A071302	Inappropriate/Inadequate Shock/Stimulation		
									A090801	Signal Artifact/Noise		
									A1301	Failure to Read Input Signal		
11507	2022	All other reportable incidents	35852	Single-chamber implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E2104	Electric Shock	A070909	Oversensing		

									A071302	Inappropriate/Inadequate Shock/Stimulation		
									A072201	High Impedance		
									A090801	Signal Artifact/Noise		
							F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A070909	Oversensing		
									A071302	Inappropriate/Inadequate Shock/Stimulation		
									A072201	High Impedance		
									A090801	Signal Artifact/Noise		
11508	2022	All other reportable incidents	35852	Single-chamber implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	17/08/2020	17/08/2020
							F08	Hospitalization or Prolonged Hospitalization	A24	No Apparent Device Problem	17/08/2020	17/08/2020
							F1905	Device Revision or Replacement	A24	No Apparent Device Problem	17/08/2020	17/08/2020
11509	2022	All other reportable incidents			CT257	Implantable cardiac pacemakers	E2104	Electric Shock	A070909	Oversensing		
									A071302	Inappropriate/Inadequate Shock/Stimulation		
									A072201	High Impedance		
									A090801	Signal Artifact/Noise		
							F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A070909	Oversensing		
									A071302	Inappropriate/Inadequate Shock/Stimulation		
									A072201	High Impedance		
									A090801	Signal Artifact/Noise		
11510	2022	All other reportable incidents			CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem		
							F26	No Health Consequences or Impact	A0705	Battery Problem		
11511	2022	All other reportable incidents	35852	Single-chamber implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E0601	Arrhythmia	A071301	Failure to Deliver Shock/Stimulation		
							F23	Unexpected Medical Intervention	A071301	Failure to Deliver Shock/Stimulation		
11512	2022	All other reportable incidents	37265	Dual-chamber implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E0112	Dizziness	A070909	Oversensing		
									A19	Environmental Compatibility Problem		
							E0601	Arrhythmia	A070909	Oversensing		
									A19	Environmental Compatibility Problem		
							F26	No Health Consequences or Impact	A070909	Oversensing		
									A19	Environmental Compatibility Problem		
11513	2022	All other reportable incidents	47265	Dual-chamber implantable pacemaker, rate-responsive	CT141	Electrodes/Leads and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	26/10/2022	26/10/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	26/10/2022	26/10/2022
					CT145	Cardiovascular devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	26/10/2022	26/10/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	26/10/2022	26/10/2022
					CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	26/10/2022	26/10/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	26/10/2022	26/10/2022
					CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	26/10/2022	26/10/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	26/10/2022	26/10/2022
					CT233	Surgical	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	26/10/2022	26/10/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	26/10/2022	26/10/2022
					CT2406	Implantable	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	26/10/2022	26/10/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	26/10/2022	26/10/2022
					CT2407	Active implantable	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	26/10/2022	26/10/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	26/10/2022	26/10/2022
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	26/10/2022	26/10/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	26/10/2022	26/10/2022
					CT2727	Active implant devices/ancillaries	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	26/10/2022	26/10/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	26/10/2022	26/10/2022
					CT2750	Implantable cardiac electrical stimulators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	26/10/2022	26/10/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	26/10/2022	26/10/2022
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	26/10/2022	26/10/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	26/10/2022	26/10/2022
					CT2985	Active	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	26/10/2022	26/10/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	26/10/2022	26/10/2022
					CT3001	Implantable metal	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	26/10/2022	26/10/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	26/10/2022	26/10/2022
					CT3002	Implantable synthetic polymer	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	26/10/2022	26/10/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	26/10/2022	26/10/2022
					CT3006	Implantable inorganic material	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	26/10/2022	26/10/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	26/10/2022	26/10/2022
					CT314	Electrically powered	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	26/10/2022	26/10/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	26/10/2022	26/10/2022
					CT321	Long-term surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	26/10/2022	26/10/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	26/10/2022	26/10/2022
					CT449	Generators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	26/10/2022	26/10/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	26/10/2022	26/10/2022
					CT981	Single use	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	26/10/2022	26/10/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	26/10/2022	26/10/2022
					CT983	Surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	26/10/2022	26/10/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	26/10/2022	26/10/2022
11514	2022	All other reportable incidents	47265	Dual-chamber implantable pacemaker, rate-responsive	CT141	Electrodes/Leads and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	16/10/2022	16/10/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	16/10/2022	16/10/2022
					CT145	Cardiovascular devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	16/10/2022	16/10/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	16/10/2022	16/10/2022
					CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	16/10/2022	16/10/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	16/10/2022	16/10/2022
					CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	16/10/2022	16/10/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	16/10/2022	16/10/2022
					CT233	Surgical	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	16/10/2022	16/10/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	16/10/2022	16/10/2022
					CT2406	Implantable	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	16/10/2022	16/10/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	16/10/2022	16/10/2022
					CT2407	Active implantable	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	16/10/2022	16/10/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	16/10/2022	16/10/2022
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	16/10/2022	16/10/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	16/10/2022	16/10/2022
					CT2727	Active implant devices/ancillaries	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	16/10/2022	16/10/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	16/10/2022	16/10/2022
					CT2750	Implantable cardiac electrical stimulators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	16/10/2022	16/10/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	16/10/2022	16/10/2022

					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	16/10/2022	16/10/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	16/10/2022	16/10/2022
					CT2985	Active	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	16/10/2022	16/10/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	16/10/2022	16/10/2022
					CT3001	Implantable metal	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	16/10/2022	16/10/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	16/10/2022	16/10/2022
					CT3002	Implantable synthetic polymer	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	16/10/2022	16/10/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	16/10/2022	16/10/2022
					CT3006	Implantable inorganic material	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	16/10/2022	16/10/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	16/10/2022	16/10/2022
					CT314	Electrically powered	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	16/10/2022	16/10/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	16/10/2022	16/10/2022
					CT321	Long-term surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	16/10/2022	16/10/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	16/10/2022	16/10/2022
					CT449	Generators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	16/10/2022	16/10/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	16/10/2022	16/10/2022
					CT981	Single use	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	16/10/2022	16/10/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	16/10/2022	16/10/2022
					CT983	Surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	16/10/2022	16/10/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	16/10/2022	16/10/2022
11515	2022	All other reportable incidents	47265	Dual-chamber implantable pacemaker, rate-responsive	CT141	Electrodes/Leads and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	07/10/2022	07/10/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	07/10/2022	07/10/2022
					CT145	Cardiovascular devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	07/10/2022	07/10/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	07/10/2022	07/10/2022
					CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	07/10/2022	07/10/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	07/10/2022	07/10/2022
					CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	07/10/2022	07/10/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	07/10/2022	07/10/2022
					CT233	Surgical	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	07/10/2022	07/10/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	07/10/2022	07/10/2022
					CT2406	Implantable	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	07/10/2022	07/10/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	07/10/2022	07/10/2022
					CT2407	Active implantable	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	07/10/2022	07/10/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	07/10/2022	07/10/2022
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	07/10/2022	07/10/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	07/10/2022	07/10/2022
					CT2727	Active implant devices/ancillaries	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	07/10/2022	07/10/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	07/10/2022	07/10/2022
					CT2750	Implantable cardiac electrical stimulators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	07/10/2022	07/10/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	07/10/2022	07/10/2022
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	07/10/2022	07/10/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	07/10/2022	07/10/2022
					CT2985	Active	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	07/10/2022	07/10/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	07/10/2022	07/10/2022
					CT3001	Implantable metal	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	07/10/2022	07/10/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	07/10/2022	07/10/2022
					CT3002	Implantable synthetic polymer	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	07/10/2022	07/10/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	07/10/2022	07/10/2022
					CT3006	Implantable inorganic material	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	07/10/2022	07/10/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	07/10/2022	07/10/2022
					CT314	Electrically powered	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	07/10/2022	07/10/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	07/10/2022	07/10/2022
					CT321	Long-term surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	07/10/2022	07/10/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	07/10/2022	07/10/2022
					CT449	Generators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	07/10/2022	07/10/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	07/10/2022	07/10/2022
					CT981	Single use	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	07/10/2022	07/10/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	07/10/2022	07/10/2022
					CT983	Surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	07/10/2022	07/10/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	07/10/2022	07/10/2022
11516	2022	All other reportable incidents	37265	Dual-chamber implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
11517	2022	All other reportable incidents	17882	Physiologic-monitoring defibrillation system	CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A0902	Display or Visual Feedback Problem		
							F27	Problem Identified During Non-Clinical Procedure	A0902	Display or Visual Feedback Problem		
11518	2022	All other reportable incidents	47265	Dual-chamber implantable pacemaker, rate-responsive			E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
									A0908	Incorrect, Inadequate or Imprecise Result or Readings		
							F26	No Health Consequences or Impact	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
									A0908	Incorrect, Inadequate or Imprecise Result or Readings		
11519	2022	All other reportable incidents	35852	Single-chamber implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E2104	Electric Shock	A070909	Oversensing		
									A071302	Inappropriate/Inadequate Shock/Stimulation		
									A090801	Signal Artifact/Noise		
							F08	Hospitalization or Prolonged Hospitalization	A070909	Oversensing		
									A071302	Inappropriate/Inadequate Shock/Stimulation		
									A090801	Signal Artifact/Noise		
							F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A070909	Oversensing		
									A071302	Inappropriate/Inadequate Shock/Stimulation		
									A090801	Signal Artifact/Noise		
							F1905	Device Revision or Replacement	A070909	Oversensing		
									A071302	Inappropriate/Inadequate Shock/Stimulation		
									A090801	Signal Artifact/Noise		
							F2203	Imaging Required	A070909	Oversensing		
									A071302	Inappropriate/Inadequate Shock/Stimulation		
									A090801	Signal Artifact/Noise		
11520	2022	All other reportable incidents	36241	Pacemaker/defibrillator lead	CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070102	High Capture Threshold		

							F08	Hospitalization or Prolonged Hospitalization	A070102	High Capture Threshold		
							F1905	Device Revision or Replacement	A070102	High Capture Threshold		
11521	2022	All other reportable incidents	47263	Cardiac resynchronization therapy implantable pacemaker	CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A0908	Incorrect, Inadequate or Imprecise Result or Readings	29/11/2022	29/11/2022
									A160503	Inappropriate or Unexpected Reset	29/11/2022	29/11/2022
							F08	Hospitalization or Prolonged Hospitalization	A0908	Incorrect, Inadequate or Imprecise Result or Readings	29/11/2022	29/11/2022
									A160503	Inappropriate or Unexpected Reset	29/11/2022	29/11/2022
							F1905	Device Revision or Replacement	A0908	Incorrect, Inadequate or Imprecise Result or Readings	29/11/2022	29/11/2022
									A160503	Inappropriate or Unexpected Reset	29/11/2022	29/11/2022
					CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A0908	Incorrect, Inadequate or Imprecise Result or Readings	29/11/2022	29/11/2022
									A160503	Inappropriate or Unexpected Reset	29/11/2022	29/11/2022
							F08	Hospitalization or Prolonged Hospitalization	A0908	Incorrect, Inadequate or Imprecise Result or Readings	29/11/2022	29/11/2022
									A160503	Inappropriate or Unexpected Reset	29/11/2022	29/11/2022
							F1905	Device Revision or Replacement	A0908	Incorrect, Inadequate or Imprecise Result or Readings	29/11/2022	29/11/2022
									A160503	Inappropriate or Unexpected Reset	29/11/2022	29/11/2022
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A0908	Incorrect, Inadequate or Imprecise Result or Readings	29/11/2022	29/11/2022
									A160503	Inappropriate or Unexpected Reset	29/11/2022	29/11/2022
							F08	Hospitalization or Prolonged Hospitalization	A0908	Incorrect, Inadequate or Imprecise Result or Readings	29/11/2022	29/11/2022
									A160503	Inappropriate or Unexpected Reset	29/11/2022	29/11/2022
							F1905	Device Revision or Replacement	A0908	Incorrect, Inadequate or Imprecise Result or Readings	29/11/2022	29/11/2022
									A160503	Inappropriate or Unexpected Reset	29/11/2022	29/11/2022
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A0908	Incorrect, Inadequate or Imprecise Result or Readings	29/11/2022	29/11/2022
									A160503	Inappropriate or Unexpected Reset	29/11/2022	29/11/2022
							F08	Hospitalization or Prolonged Hospitalization	A0908	Incorrect, Inadequate or Imprecise Result or Readings	29/11/2022	29/11/2022
									A160503	Inappropriate or Unexpected Reset	29/11/2022	29/11/2022
							F1905	Device Revision or Replacement	A0908	Incorrect, Inadequate or Imprecise Result or Readings	29/11/2022	29/11/2022
									A160503	Inappropriate or Unexpected Reset	29/11/2022	29/11/2022
11522	2022	Unanticipated serious deterioration in state of health	35852	Single-chamber implantable defibrillator	CT2866	Implantable cardiac electrical stimulators			A26	Insufficient Device Problem Information		
11523	2022		47265	Dual-chamber implantable pacemaker, rate-responsive	CT141	Electrodes/Leads and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	21/10/2022	21/10/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	21/10/2022	21/10/2022
					CT145	Cardiovascular devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	21/10/2022	21/10/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	21/10/2022	21/10/2022
					CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	21/10/2022	21/10/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	21/10/2022	21/10/2022
					CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	21/10/2022	21/10/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	21/10/2022	21/10/2022
					CT233	Surgical	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	21/10/2022	21/10/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	21/10/2022	21/10/2022
					CT2406	Implantable	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	21/10/2022	21/10/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	21/10/2022	21/10/2022
					CT2407	Active implantable	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	21/10/2022	21/10/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	21/10/2022	21/10/2022
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	21/10/2022	21/10/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	21/10/2022	21/10/2022
					CT2727	Active implant devices/ancillaries	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	21/10/2022	21/10/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	21/10/2022	21/10/2022
					CT2750	Implantable cardiac electrical stimulators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	21/10/2022	21/10/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	21/10/2022	21/10/2022
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	21/10/2022	21/10/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	21/10/2022	21/10/2022
					CT2985	Active	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	21/10/2022	21/10/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	21/10/2022	21/10/2022
					CT3001	Implantable metal	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	21/10/2022	21/10/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	21/10/2022	21/10/2022
					CT3002	Implantable synthetic polymer	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	21/10/2022	21/10/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	21/10/2022	21/10/2022
					CT3006	Implantable inorganic material	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	21/10/2022	21/10/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	21/10/2022	21/10/2022
					CT314	Electrically powered	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	21/10/2022	21/10/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	21/10/2022	21/10/2022
					CT321	Long-term surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	21/10/2022	21/10/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	21/10/2022	21/10/2022
					CT449	Generators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	21/10/2022	21/10/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	21/10/2022	21/10/2022
					CT981	Single use	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	21/10/2022	21/10/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	21/10/2022	21/10/2022
					CT983	Surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	21/10/2022	21/10/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	21/10/2022	21/10/2022
11524	2022	All other reportable incidents	47265	Dual-chamber implantable pacemaker, rate-responsive	CT141	Electrodes/Leads and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A071102	Failure to Interrogate		
									A071204	No Pacing		
							F26	No Health Consequences or Impact	A071102	Failure to Interrogate		
									A071204	No Pacing		
					CT145	Cardiovascular devices	E2403	No Clinical Signs, Symptoms or Conditions	A071102	Failure to Interrogate		
									A071204	No Pacing		
							F26	No Health Consequences or Impact	A071102	Failure to Interrogate		
									A071204	No Pacing		
					CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A071102	Failure to Interrogate		
									A071204	No Pacing		
							F26	No Health Consequences or Impact	A071102	Failure to Interrogate		

									A071204	No Pacing		
					CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A071102	Failure to Interrogate		
									A071204	No Pacing		
							F26	No Health Consequences or Impact	A071102	Failure to Interrogate		
									A071204	No Pacing		
					CT233	Surgical	E2403	No Clinical Signs, Symptoms or Conditions	A071102	Failure to Interrogate		
									A071204	No Pacing		
							F26	No Health Consequences or Impact	A071102	Failure to Interrogate		
									A071204	No Pacing		
					CT2406	Implantable	E2403	No Clinical Signs, Symptoms or Conditions	A071102	Failure to Interrogate		
									A071204	No Pacing		
							F26	No Health Consequences or Impact	A071102	Failure to Interrogate		
									A071204	No Pacing		
					CT2407	Active implantable	E2403	No Clinical Signs, Symptoms or Conditions	A071102	Failure to Interrogate		
									A071204	No Pacing		
							F26	No Health Consequences or Impact	A071102	Failure to Interrogate		
									A071204	No Pacing		
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A071102	Failure to Interrogate		
									A071204	No Pacing		
							F26	No Health Consequences or Impact	A071102	Failure to Interrogate		
									A071204	No Pacing		
					CT2727	Active implant devices/ancillaries	E2403	No Clinical Signs, Symptoms or Conditions	A071102	Failure to Interrogate		
									A071204	No Pacing		
							F26	No Health Consequences or Impact	A071102	Failure to Interrogate		
									A071204	No Pacing		
					CT2750	Implantable cardiac electrical stimulators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A071102	Failure to Interrogate		
									A071204	No Pacing		
							F26	No Health Consequences or Impact	A071102	Failure to Interrogate		
									A071204	No Pacing		
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A071102	Failure to Interrogate		
									A071204	No Pacing		
							F26	No Health Consequences or Impact	A071102	Failure to Interrogate		
									A071204	No Pacing		
					CT2985	Active	E2403	No Clinical Signs, Symptoms or Conditions	A071102	Failure to Interrogate		
									A071204	No Pacing		
							F26	No Health Consequences or Impact	A071102	Failure to Interrogate		
									A071204	No Pacing		
					CT3001	Implantable metal	E2403	No Clinical Signs, Symptoms or Conditions	A071102	Failure to Interrogate		
									A071204	No Pacing		
							F26	No Health Consequences or Impact	A071102	Failure to Interrogate		
									A071204	No Pacing		
					CT3002	Implantable synthetic polymer	E2403	No Clinical Signs, Symptoms or Conditions	A071102	Failure to Interrogate		
									A071204	No Pacing		
							F26	No Health Consequences or Impact	A071102	Failure to Interrogate		
									A071204	No Pacing		
					CT3006	Implantable inorganic material	E2403	No Clinical Signs, Symptoms or Conditions	A071102	Failure to Interrogate		
									A071204	No Pacing		
							F26	No Health Consequences or Impact	A071102	Failure to Interrogate		
									A071204	No Pacing		
					CT314	Electrically powered	E2403	No Clinical Signs, Symptoms or Conditions	A071102	Failure to Interrogate		
									A071204	No Pacing		
							F26	No Health Consequences or Impact	A071102	Failure to Interrogate		
									A071204	No Pacing		
					CT321	Long-term surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A071102	Failure to Interrogate		
									A071204	No Pacing		
							F26	No Health Consequences or Impact	A071102	Failure to Interrogate		
									A071204	No Pacing		
					CT449	Generators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A071102	Failure to Interrogate		
									A071204	No Pacing		
							F26	No Health Consequences or Impact	A071102	Failure to Interrogate		
									A071204	No Pacing		
					CT981	Single use	E2403	No Clinical Signs, Symptoms or Conditions	A071102	Failure to Interrogate		
									A071204	No Pacing		
							F26	No Health Consequences or Impact	A071102	Failure to Interrogate		
									A071204	No Pacing		
					CT983	Surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A071102	Failure to Interrogate		
									A071204	No Pacing		
							F26	No Health Consequences or Impact	A071102	Failure to Interrogate		
									A071204	No Pacing		
11525	2022		47265	Dual-chamber implantable pacemaker, rate-responsive	CT141	Electrodes/Leads and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	08/10/2022	08/10/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	08/10/2022	08/10/2022
					CT145	Cardiovascular devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	08/10/2022	08/10/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	08/10/2022	08/10/2022
					CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	08/10/2022	08/10/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	08/10/2022	08/10/2022
					CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	08/10/2022	08/10/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	08/10/2022	08/10/2022
					CT233	Surgical	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	08/10/2022	08/10/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	08/10/2022	08/10/2022
					CT2406	Implantable	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	08/10/2022	08/10/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	08/10/2022	08/10/2022
					CT2407	Active implantable	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	08/10/2022	08/10/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	08/10/2022	08/10/2022
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	08/10/2022	08/10/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	08/10/2022	08/10/2022
					CT2727	Active implant devices/ancillaries	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	08/10/2022	08/10/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	08/10/2022	08/10/2022
					CT2750	Implantable cardiac electrical stimulators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	08/10/2022	08/10/2022

							F1901	Additional Surgery	A24	No Apparent Device Problem	08/10/2022	08/10/2022
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	08/10/2022	08/10/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	08/10/2022	08/10/2022
					CT2985	Active	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	08/10/2022	08/10/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	08/10/2022	08/10/2022
					CT3001	Implantable metal	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	08/10/2022	08/10/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	08/10/2022	08/10/2022
					CT3002	Implantable synthetic polymer	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	08/10/2022	08/10/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	08/10/2022	08/10/2022
					CT3006	Implantable inorganic material	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	08/10/2022	08/10/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	08/10/2022	08/10/2022
					CT314	Electrically powered	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	08/10/2022	08/10/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	08/10/2022	08/10/2022
					CT321	Long-term surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	08/10/2022	08/10/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	08/10/2022	08/10/2022
					CT449	Generators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	08/10/2022	08/10/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	08/10/2022	08/10/2022
					CT981	Single use	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	08/10/2022	08/10/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	08/10/2022	08/10/2022
					CT983	Surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	08/10/2022	08/10/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	08/10/2022	08/10/2022
11526	2022	Unanticipated serious deterioration in state of health	60789	Intracardiac pacemaker	CT130	Catheters and associated devices	F	Annex F	A0101	Patient-Device Incompatibility		
					CT145	Cardiovascular devices	F	Annex F	A0101	Patient-Device Incompatibility		
					CT1684	Cardiac pacemakers and associated devices	F	Annex F	A0101	Patient-Device Incompatibility		
					CT1685	Cardiac pacemakers	F	Annex F	A0101	Patient-Device Incompatibility		
					CT233	Surgical	F	Annex F	A0101	Patient-Device Incompatibility		
					CT2406	Implantable	F	Annex F	A0101	Patient-Device Incompatibility		
					CT2407	Active implantable	F	Annex F	A0101	Patient-Device Incompatibility		
					CT257	Implantable cardiac pacemakers	F	Annex F	A0101	Patient-Device Incompatibility		
					CT2727	Active implant devices/ancillaries	F	Annex F	A0101	Patient-Device Incompatibility		
					CT2750	Implantable cardiac electrical stimulators and associated devices	F	Annex F	A0101	Patient-Device Incompatibility		
					CT2818	Catheterization/Cannulation/Fluid path devices	F	Annex F	A0101	Patient-Device Incompatibility		
					CT2866	Implantable cardiac electrical stimulators	F	Annex F	A0101	Patient-Device Incompatibility		
					CT2985	Active	F	Annex F	A0101	Patient-Device Incompatibility		
					CT3001	Implantable metal	F	Annex F	A0101	Patient-Device Incompatibility		
					CT3002	Implantable synthetic polymer	F	Annex F	A0101	Patient-Device Incompatibility		
					CT3006	Implantable inorganic material	F	Annex F	A0101	Patient-Device Incompatibility		
					CT314	Electrically powered	F	Annex F	A0101	Patient-Device Incompatibility		
					CT321	Long-term surgical invasive	F	Annex F	A0101	Patient-Device Incompatibility		
					CT981	Single use	F	Annex F	A0101	Patient-Device Incompatibility		
					CT983	Surgical invasive	F	Annex F	A0101	Patient-Device Incompatibility		
11527	2022	All other reportable incidents	47263	Cardiac resynchronization therapy implantable pacemaker	CT1684	Cardiac pacemakers and associated devices	E060106	Heart Block	A01	Patient Device Interaction Problem	24/10/2022	
							E2403	No Clinical Signs, Symptoms or Conditions	A01	Patient Device Interaction Problem	24/10/2022	
							F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A01	Patient Device Interaction Problem	24/10/2022	
					CT1685	Cardiac pacemakers	E060106	Heart Block	A01	Patient Device Interaction Problem	24/10/2022	
							E2403	No Clinical Signs, Symptoms or Conditions	A01	Patient Device Interaction Problem	24/10/2022	
							F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A01	Patient Device Interaction Problem	24/10/2022	
					CT257	Implantable cardiac pacemakers	E060106	Heart Block	A01	Patient Device Interaction Problem	24/10/2022	
							E2403	No Clinical Signs, Symptoms or Conditions	A01	Patient Device Interaction Problem	24/10/2022	
							F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A01	Patient Device Interaction Problem	24/10/2022	
					CT2866	Implantable cardiac electrical stimulators	E060106	Heart Block	A01	Patient Device Interaction Problem	24/10/2022	
							E2403	No Clinical Signs, Symptoms or Conditions	A01	Patient Device Interaction Problem	24/10/2022	
							F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A01	Patient Device Interaction Problem	24/10/2022	
11528	2022	All other reportable incidents	47270	Cardiac resynchronization therapy implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A0712	Pacing Problem		
							F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A0712	Pacing Problem		
11529	2022	All other reportable incidents	35852	Single-chamber implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E2104	Electric Shock	A0709	Device Sensing Problem	01/12/2022	01/12/2022
									A070909	Oversensing	01/12/2022	01/12/2022
									A071302	Inappropriate/Inadequate Shock/Stimulation	01/12/2022	01/12/2022
									A090801	Signal Artifact/Noise	01/12/2022	01/12/2022
							F08	Hospitalization or Prolonged Hospitalization	A0709	Device Sensing Problem	01/12/2022	01/12/2022
									A070909	Oversensing	01/12/2022	01/12/2022
									A071302	Inappropriate/Inadequate Shock/Stimulation	01/12/2022	01/12/2022
									A090801	Signal Artifact/Noise	01/12/2022	01/12/2022
							F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A0709	Device Sensing Problem	01/12/2022	01/12/2022
									A070909	Oversensing	01/12/2022	01/12/2022
									A071302	Inappropriate/Inadequate Shock/Stimulation	01/12/2022	01/12/2022
									A090801	Signal Artifact/Noise	01/12/2022	01/12/2022
							F1903	Device Explantation	A0709	Device Sensing Problem	01/12/2022	01/12/2022
									A070909	Oversensing	01/12/2022	01/12/2022
									A071302	Inappropriate/Inadequate Shock/Stimulation	01/12/2022	01/12/2022
									A090801	Signal Artifact/Noise	01/12/2022	01/12/2022
							F2203	Imaging Required	A0709	Device Sensing Problem	01/12/2022	01/12/2022
									A070909	Oversensing	01/12/2022	01/12/2022
									A071302	Inappropriate/Inadequate Shock/Stimulation	01/12/2022	01/12/2022
									A090801	Signal Artifact/Noise	01/12/2022	01/12/2022
							F23	Unexpected Medical Intervention	A0709	Device Sensing Problem	01/12/2022	01/12/2022
									A070909	Oversensing	01/12/2022	01/12/2022
									A071302	Inappropriate/Inadequate Shock/Stimulation	01/12/2022	01/12/2022
									A090801	Signal Artifact/Noise	01/12/2022	01/12/2022
11530	2022	All other reportable incidents	47265	Dual-chamber implantable pacemaker, rate-responsive	CT141	Electrodes/Leads and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
					CT145	Cardiovascular devices	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		

					CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
					CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
					CT233	Surgical	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
					CT2406	Implantable	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
					CT2407	Active implantable	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
					CT2727	Active implant devices/ancillaries	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
					CT2750	Implantable cardiac electrical stimulators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
					CT2985	Active	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
					CT3001	Implantable metal	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
					CT3002	Implantable synthetic polymer	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
					CT3006	Implantable inorganic material	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
					CT314	Electrically powered	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
					CT321	Long-term surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
					CT449	Generators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
					CT981	Single use	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
					CT983	Surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
11531	2022	All other reportable incidents	47265	Dual-chamber implantable pacemaker, rate-responsive	CT141	Electrodes/Leads and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070909	Oversensing	30/09/2022	30/09/2022
									A090801	Signal Artifact/Noise	30/09/2022	30/09/2022
									A180104	Device Contamination with Chemical or Other Material	30/09/2022	30/09/2022
							F08	Hospitalization or Prolonged Hospitalization	A070909	Oversensing	30/09/2022	30/09/2022
									A090801	Signal Artifact/Noise	30/09/2022	30/09/2022
									A180104	Device Contamination with Chemical or Other Material	30/09/2022	30/09/2022
							F1905	Device Revision or Replacement	A070909	Oversensing	30/09/2022	30/09/2022
									A090801	Signal Artifact/Noise	30/09/2022	30/09/2022
									A180104	Device Contamination with Chemical or Other Material	30/09/2022	30/09/2022
							F2203	Imaging Required	A070909	Oversensing	30/09/2022	30/09/2022
									A090801	Signal Artifact/Noise	30/09/2022	30/09/2022
									A180104	Device Contamination with Chemical or Other Material	30/09/2022	30/09/2022
					CT145	Cardiovascular devices	E2403	No Clinical Signs, Symptoms or Conditions	A070909	Oversensing	30/09/2022	30/09/2022

								A090801	Signal Artifact/Noise	30/09/2022	30/09/2022
								A180104	Device Contamination with Chemical or Other Material	30/09/2022	30/09/2022
						F08	Hospitalization or Prolonged Hospitalization	A070909	Oversensing	30/09/2022	30/09/2022
								A090801	Signal Artifact/Noise	30/09/2022	30/09/2022
								A180104	Device Contamination with Chemical or Other Material	30/09/2022	30/09/2022
						F1905	Device Revision or Replacement	A070909	Oversensing	30/09/2022	30/09/2022
								A090801	Signal Artifact/Noise	30/09/2022	30/09/2022
								A180104	Device Contamination with Chemical or Other Material	30/09/2022	30/09/2022
						F2203	Imaging Required	A070909	Oversensing	30/09/2022	30/09/2022
								A090801	Signal Artifact/Noise	30/09/2022	30/09/2022
								A180104	Device Contamination with Chemical or Other Material	30/09/2022	30/09/2022
					CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070909	Oversensing	30/09/2022	30/09/2022
									A090801	Signal Artifact/Noise	30/09/2022	30/09/2022
									A180104	Device Contamination with Chemical or Other Material	30/09/2022	30/09/2022
							F08	Hospitalization or Prolonged Hospitalization	A070909	Oversensing	30/09/2022	30/09/2022
									A090801	Signal Artifact/Noise	30/09/2022	30/09/2022
									A180104	Device Contamination with Chemical or Other Material	30/09/2022	30/09/2022
							F1905	Device Revision or Replacement	A070909	Oversensing	30/09/2022	30/09/2022
									A090801	Signal Artifact/Noise	30/09/2022	30/09/2022
									A180104	Device Contamination with Chemical or Other Material	30/09/2022	30/09/2022
							F2203	Imaging Required	A070909	Oversensing	30/09/2022	30/09/2022
									A090801	Signal Artifact/Noise	30/09/2022	30/09/2022
									A180104	Device Contamination with Chemical or Other Material	30/09/2022	30/09/2022
					CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A070909	Oversensing	30/09/2022	30/09/2022
									A090801	Signal Artifact/Noise	30/09/2022	30/09/2022
									A180104	Device Contamination with Chemical or Other Material	30/09/2022	30/09/2022
							F08	Hospitalization or Prolonged Hospitalization	A070909	Oversensing	30/09/2022	30/09/2022
									A090801	Signal Artifact/Noise	30/09/2022	30/09/2022
									A180104	Device Contamination with Chemical or Other Material	30/09/2022	30/09/2022
							F1905	Device Revision or Replacement	A070909	Oversensing	30/09/2022	30/09/2022
									A090801	Signal Artifact/Noise	30/09/2022	30/09/2022
									A180104	Device Contamination with Chemical or Other Material	30/09/2022	30/09/2022
							F2203	Imaging Required	A070909	Oversensing	30/09/2022	30/09/2022
									A090801	Signal Artifact/Noise	30/09/2022	30/09/2022
									A180104	Device Contamination with Chemical or Other Material	30/09/2022	30/09/2022
					CT233	Surgical	E2403	No Clinical Signs, Symptoms or Conditions	A070909	Oversensing	30/09/2022	30/09/2022
									A090801	Signal Artifact/Noise	30/09/2022	30/09/2022
									A180104	Device Contamination with Chemical or Other Material	30/09/2022	30/09/2022
							F08	Hospitalization or Prolonged Hospitalization	A070909	Oversensing	30/09/2022	30/09/2022
									A090801	Signal Artifact/Noise	30/09/2022	30/09/2022
									A180104	Device Contamination with Chemical or Other Material	30/09/2022	30/09/2022
							F1905	Device Revision or Replacement	A070909	Oversensing	30/09/2022	30/09/2022
									A090801	Signal Artifact/Noise	30/09/2022	30/09/2022
									A180104	Device Contamination with Chemical or Other Material	30/09/2022	30/09/2022
							F2203	Imaging Required	A070909	Oversensing	30/09/2022	30/09/2022
									A090801	Signal Artifact/Noise	30/09/2022	30/09/2022
									A180104	Device Contamination with Chemical or Other Material	30/09/2022	30/09/2022
					CT2406	Implantable	E2403	No Clinical Signs, Symptoms or Conditions	A070909	Oversensing	30/09/2022	30/09/2022
									A090801	Signal Artifact/Noise	30/09/2022	30/09/2022
									A180104	Device Contamination with Chemical or Other Material	30/09/2022	30/09/2022
							F08	Hospitalization or Prolonged Hospitalization	A070909	Oversensing	30/09/2022	30/09/2022
									A090801	Signal Artifact/Noise	30/09/2022	30/09/2022
									A180104	Device Contamination with Chemical or Other Material	30/09/2022	30/09/2022
							F1905	Device Revision or Replacement	A070909	Oversensing	30/09/2022	30/09/2022
									A090801	Signal Artifact/Noise	30/09/2022	30/09/2022
									A180104	Device Contamination with Chemical or Other Material	30/09/2022	30/09/2022
							F2203	Imaging Required	A070909	Oversensing	30/09/2022	30/09/2022
									A090801	Signal Artifact/Noise	30/09/2022	30/09/2022
									A180104	Device Contamination with Chemical or Other Material	30/09/2022	30/09/2022
					CT2407	Active implantable	E2403	No Clinical Signs, Symptoms or Conditions	A070909	Oversensing	30/09/2022	30/09/2022
									A090801	Signal Artifact/Noise	30/09/2022	30/09/2022
									A180104	Device Contamination with Chemical or Other Material	30/09/2022	30/09/2022
							F08	Hospitalization or Prolonged Hospitalization	A070909	Oversensing	30/09/2022	30/09/2022
									A090801	Signal Artifact/Noise	30/09/2022	30/09/2022
									A180104	Device Contamination with Chemical or Other Material	30/09/2022	30/09/2022
							F1905	Device Revision or Replacement	A070909	Oversensing	30/09/2022	30/09/2022
									A090801	Signal Artifact/Noise	30/09/2022	30/09/2022
									A180104	Device Contamination with Chemical or Other Material	30/09/2022	30/09/2022
							F2203	Imaging Required	A070909	Oversensing	30/09/2022	30/09/2022
									A090801	Signal Artifact/Noise	30/09/2022	30/09/2022

									A180104	Device Contamination with Chemical or Other Material	30/09/2022	30/09/2022
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A070909	Oversensing	30/09/2022	30/09/2022
									A090801	Signal Artifact/Noise	30/09/2022	30/09/2022
									A180104	Device Contamination with Chemical or Other Material	30/09/2022	30/09/2022
							F08	Hospitalization or Prolonged Hospitalization	A070909	Oversensing	30/09/2022	30/09/2022
									A090801	Signal Artifact/Noise	30/09/2022	30/09/2022
									A180104	Device Contamination with Chemical or Other Material	30/09/2022	30/09/2022
							F1905	Device Revision or Replacement	A070909	Oversensing	30/09/2022	30/09/2022
									A090801	Signal Artifact/Noise	30/09/2022	30/09/2022
									A180104	Device Contamination with Chemical or Other Material	30/09/2022	30/09/2022
							F2203	Imaging Required	A070909	Oversensing	30/09/2022	30/09/2022
									A090801	Signal Artifact/Noise	30/09/2022	30/09/2022
									A180104	Device Contamination with Chemical or Other Material	30/09/2022	30/09/2022
					CT2727	Active implant devices/ancillaries	E2403	No Clinical Signs, Symptoms or Conditions	A070909	Oversensing	30/09/2022	30/09/2022
									A090801	Signal Artifact/Noise	30/09/2022	30/09/2022
									A180104	Device Contamination with Chemical or Other Material	30/09/2022	30/09/2022
							F08	Hospitalization or Prolonged Hospitalization	A070909	Oversensing	30/09/2022	30/09/2022
									A090801	Signal Artifact/Noise	30/09/2022	30/09/2022
									A180104	Device Contamination with Chemical or Other Material	30/09/2022	30/09/2022
							F1905	Device Revision or Replacement	A070909	Oversensing	30/09/2022	30/09/2022
									A090801	Signal Artifact/Noise	30/09/2022	30/09/2022
									A180104	Device Contamination with Chemical or Other Material	30/09/2022	30/09/2022
							F2203	Imaging Required	A070909	Oversensing	30/09/2022	30/09/2022
									A090801	Signal Artifact/Noise	30/09/2022	30/09/2022
									A180104	Device Contamination with Chemical or Other Material	30/09/2022	30/09/2022
					CT2750	Implantable cardiac electrical stimulators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070909	Oversensing	30/09/2022	30/09/2022
									A090801	Signal Artifact/Noise	30/09/2022	30/09/2022
									A180104	Device Contamination with Chemical or Other Material	30/09/2022	30/09/2022
							F08	Hospitalization or Prolonged Hospitalization	A070909	Oversensing	30/09/2022	30/09/2022
									A090801	Signal Artifact/Noise	30/09/2022	30/09/2022
									A180104	Device Contamination with Chemical or Other Material	30/09/2022	30/09/2022
							F1905	Device Revision or Replacement	A070909	Oversensing	30/09/2022	30/09/2022
									A090801	Signal Artifact/Noise	30/09/2022	30/09/2022
									A180104	Device Contamination with Chemical or Other Material	30/09/2022	30/09/2022
							F2203	Imaging Required	A070909	Oversensing	30/09/2022	30/09/2022
									A090801	Signal Artifact/Noise	30/09/2022	30/09/2022
									A180104	Device Contamination with Chemical or Other Material	30/09/2022	30/09/2022
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A070909	Oversensing	30/09/2022	30/09/2022
									A090801	Signal Artifact/Noise	30/09/2022	30/09/2022
									A180104	Device Contamination with Chemical or Other Material	30/09/2022	30/09/2022
							F08	Hospitalization or Prolonged Hospitalization	A070909	Oversensing	30/09/2022	30/09/2022
									A090801	Signal Artifact/Noise	30/09/2022	30/09/2022
									A180104	Device Contamination with Chemical or Other Material	30/09/2022	30/09/2022
							F1905	Device Revision or Replacement	A070909	Oversensing	30/09/2022	30/09/2022
									A090801	Signal Artifact/Noise	30/09/2022	30/09/2022
									A180104	Device Contamination with Chemical or Other Material	30/09/2022	30/09/2022
							F2203	Imaging Required	A070909	Oversensing	30/09/2022	30/09/2022
									A090801	Signal Artifact/Noise	30/09/2022	30/09/2022
									A180104	Device Contamination with Chemical or Other Material	30/09/2022	30/09/2022
					CT2985	Active	E2403	No Clinical Signs, Symptoms or Conditions	A070909	Oversensing	30/09/2022	30/09/2022
									A090801	Signal Artifact/Noise	30/09/2022	30/09/2022
									A180104	Device Contamination with Chemical or Other Material	30/09/2022	30/09/2022
							F08	Hospitalization or Prolonged Hospitalization	A070909	Oversensing	30/09/2022	30/09/2022
									A090801	Signal Artifact/Noise	30/09/2022	30/09/2022
									A180104	Device Contamination with Chemical or Other Material	30/09/2022	30/09/2022
							F1905	Device Revision or Replacement	A070909	Oversensing	30/09/2022	30/09/2022
									A090801	Signal Artifact/Noise	30/09/2022	30/09/2022
									A180104	Device Contamination with Chemical or Other Material	30/09/2022	30/09/2022
							F2203	Imaging Required	A070909	Oversensing	30/09/2022	30/09/2022
									A090801	Signal Artifact/Noise	30/09/2022	30/09/2022
									A180104	Device Contamination with Chemical or Other Material	30/09/2022	30/09/2022
					CT3001	Implantable metal	E2403	No Clinical Signs, Symptoms or Conditions	A070909	Oversensing	30/09/2022	30/09/2022
									A090801	Signal Artifact/Noise	30/09/2022	30/09/2022
									A180104	Device Contamination with Chemical or Other Material	30/09/2022	30/09/2022
							F08	Hospitalization or Prolonged Hospitalization	A070909	Oversensing	30/09/2022	30/09/2022
									A090801	Signal Artifact/Noise	30/09/2022	30/09/2022
									A180104	Device Contamination with Chemical or Other Material	30/09/2022	30/09/2022
							F1905	Device Revision or Replacement	A070909	Oversensing	30/09/2022	30/09/2022
									A090801	Signal Artifact/Noise	30/09/2022	30/09/2022

									A180104	Device Contamination with Chemical or Other Material	30/09/2022	30/09/2022
							F2203	Imaging Required	A070909	Oversensing	30/09/2022	30/09/2022
									A090801	Signal Artifact/Noise	30/09/2022	30/09/2022
									A180104	Device Contamination with Chemical or Other Material	30/09/2022	30/09/2022
					CT3002	Implantable synthetic polymer	E2403	No Clinical Signs, Symptoms or Conditions	A070909	Oversensing	30/09/2022	30/09/2022
									A090801	Signal Artifact/Noise	30/09/2022	30/09/2022
									A180104	Device Contamination with Chemical or Other Material	30/09/2022	30/09/2022
							F08	Hospitalization or Prolonged Hospitalization	A070909	Oversensing	30/09/2022	30/09/2022
									A090801	Signal Artifact/Noise	30/09/2022	30/09/2022
									A180104	Device Contamination with Chemical or Other Material	30/09/2022	30/09/2022
							F1905	Device Revision or Replacement	A070909	Oversensing	30/09/2022	30/09/2022
									A090801	Signal Artifact/Noise	30/09/2022	30/09/2022
									A180104	Device Contamination with Chemical or Other Material	30/09/2022	30/09/2022
							F2203	Imaging Required	A070909	Oversensing	30/09/2022	30/09/2022
									A090801	Signal Artifact/Noise	30/09/2022	30/09/2022
									A180104	Device Contamination with Chemical or Other Material	30/09/2022	30/09/2022
					CT3006	Implantable inorganic material	E2403	No Clinical Signs, Symptoms or Conditions	A070909	Oversensing	30/09/2022	30/09/2022
									A090801	Signal Artifact/Noise	30/09/2022	30/09/2022
									A180104	Device Contamination with Chemical or Other Material	30/09/2022	30/09/2022
							F08	Hospitalization or Prolonged Hospitalization	A070909	Oversensing	30/09/2022	30/09/2022
									A090801	Signal Artifact/Noise	30/09/2022	30/09/2022
									A180104	Device Contamination with Chemical or Other Material	30/09/2022	30/09/2022
							F1905	Device Revision or Replacement	A070909	Oversensing	30/09/2022	30/09/2022
									A090801	Signal Artifact/Noise	30/09/2022	30/09/2022
									A180104	Device Contamination with Chemical or Other Material	30/09/2022	30/09/2022
							F2203	Imaging Required	A070909	Oversensing	30/09/2022	30/09/2022
									A090801	Signal Artifact/Noise	30/09/2022	30/09/2022
									A180104	Device Contamination with Chemical or Other Material	30/09/2022	30/09/2022
					CT314	Electrically powered	E2403	No Clinical Signs, Symptoms or Conditions	A070909	Oversensing	30/09/2022	30/09/2022
									A090801	Signal Artifact/Noise	30/09/2022	30/09/2022
									A180104	Device Contamination with Chemical or Other Material	30/09/2022	30/09/2022
							F08	Hospitalization or Prolonged Hospitalization	A070909	Oversensing	30/09/2022	30/09/2022
									A090801	Signal Artifact/Noise	30/09/2022	30/09/2022
									A180104	Device Contamination with Chemical or Other Material	30/09/2022	30/09/2022
							F1905	Device Revision or Replacement	A070909	Oversensing	30/09/2022	30/09/2022
									A090801	Signal Artifact/Noise	30/09/2022	30/09/2022
									A180104	Device Contamination with Chemical or Other Material	30/09/2022	30/09/2022
							F2203	Imaging Required	A070909	Oversensing	30/09/2022	30/09/2022
									A090801	Signal Artifact/Noise	30/09/2022	30/09/2022
									A180104	Device Contamination with Chemical or Other Material	30/09/2022	30/09/2022
					CT321	Long-term surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A070909	Oversensing	30/09/2022	30/09/2022
									A090801	Signal Artifact/Noise	30/09/2022	30/09/2022
									A180104	Device Contamination with Chemical or Other Material	30/09/2022	30/09/2022
							F08	Hospitalization or Prolonged Hospitalization	A070909	Oversensing	30/09/2022	30/09/2022
									A090801	Signal Artifact/Noise	30/09/2022	30/09/2022
									A180104	Device Contamination with Chemical or Other Material	30/09/2022	30/09/2022
							F1905	Device Revision or Replacement	A070909	Oversensing	30/09/2022	30/09/2022
									A090801	Signal Artifact/Noise	30/09/2022	30/09/2022
									A180104	Device Contamination with Chemical or Other Material	30/09/2022	30/09/2022
							F2203	Imaging Required	A070909	Oversensing	30/09/2022	30/09/2022
									A090801	Signal Artifact/Noise	30/09/2022	30/09/2022
									A180104	Device Contamination with Chemical or Other Material	30/09/2022	30/09/2022
					CT449	Generators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070909	Oversensing	30/09/2022	30/09/2022
									A090801	Signal Artifact/Noise	30/09/2022	30/09/2022
									A180104	Device Contamination with Chemical or Other Material	30/09/2022	30/09/2022
							F08	Hospitalization or Prolonged Hospitalization	A070909	Oversensing	30/09/2022	30/09/2022
									A090801	Signal Artifact/Noise	30/09/2022	30/09/2022
									A180104	Device Contamination with Chemical or Other Material	30/09/2022	30/09/2022
							F1905	Device Revision or Replacement	A070909	Oversensing	30/09/2022	30/09/2022
									A090801	Signal Artifact/Noise	30/09/2022	30/09/2022
									A180104	Device Contamination with Chemical or Other Material	30/09/2022	30/09/2022
							F2203	Imaging Required	A070909	Oversensing	30/09/2022	30/09/2022
									A090801	Signal Artifact/Noise	30/09/2022	30/09/2022
									A180104	Device Contamination with Chemical or Other Material	30/09/2022	30/09/2022
					CT981	Single use	E2403	No Clinical Signs, Symptoms or Conditions	A070909	Oversensing	30/09/2022	30/09/2022
									A090801	Signal Artifact/Noise	30/09/2022	30/09/2022
									A180104	Device Contamination with Chemical or Other Material	30/09/2022	30/09/2022
							F08	Hospitalization or Prolonged Hospitalization	A070909	Oversensing	30/09/2022	30/09/2022
									A090801	Signal Artifact/Noise	30/09/2022	30/09/2022

								A180104	Device Contamination with Chemical or Other Material	30/09/2022	30/09/2022
						F1905	Device Revision or Replacement	A070909	Oversensing	30/09/2022	30/09/2022
								A090801	Signal Artifact/Noise	30/09/2022	30/09/2022
								A180104	Device Contamination with Chemical or Other Material	30/09/2022	30/09/2022
						F2203	Imaging Required	A070909	Oversensing	30/09/2022	30/09/2022
								A090801	Signal Artifact/Noise	30/09/2022	30/09/2022
								A180104	Device Contamination with Chemical or Other Material	30/09/2022	30/09/2022
				CT983	Surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A070909	Oversensing	30/09/2022	30/09/2022
								A090801	Signal Artifact/Noise	30/09/2022	30/09/2022
								A180104	Device Contamination with Chemical or Other Material	30/09/2022	30/09/2022
						F08	Hospitalization or Prolonged Hospitalization	A070909	Oversensing	30/09/2022	30/09/2022
								A090801	Signal Artifact/Noise	30/09/2022	30/09/2022
								A180104	Device Contamination with Chemical or Other Material	30/09/2022	30/09/2022
						F1905	Device Revision or Replacement	A070909	Oversensing	30/09/2022	30/09/2022
								A090801	Signal Artifact/Noise	30/09/2022	30/09/2022
								A180104	Device Contamination with Chemical or Other Material	30/09/2022	30/09/2022
						F2203	Imaging Required	A070909	Oversensing	30/09/2022	30/09/2022
								A090801	Signal Artifact/Noise	30/09/2022	30/09/2022
								A180104	Device Contamination with Chemical or Other Material	30/09/2022	30/09/2022
11532	2022		47265	Dual-chamber implantable pacemaker, rate-responsive	CT141	Electrodes/Leads and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	06/11/2022	06/11/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	06/11/2022	06/11/2022
					CT145	Cardiovascular devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	06/11/2022	06/11/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	06/11/2022	06/11/2022
					CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	06/11/2022	06/11/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	06/11/2022	06/11/2022
					CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	06/11/2022	06/11/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	06/11/2022	06/11/2022
					CT233	Surgical	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	06/11/2022	06/11/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	06/11/2022	06/11/2022
					CT2406	Implantable	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	06/11/2022	06/11/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	06/11/2022	06/11/2022
					CT2407	Active implantable	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	06/11/2022	06/11/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	06/11/2022	06/11/2022
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	06/11/2022	06/11/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	06/11/2022	06/11/2022
					CT2727	Active implant devices/ancillaries	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	06/11/2022	06/11/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	06/11/2022	06/11/2022
					CT2750	Implantable cardiac electrical stimulators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	06/11/2022	06/11/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	06/11/2022	06/11/2022
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	06/11/2022	06/11/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	06/11/2022	06/11/2022
					CT2985	Active	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	06/11/2022	06/11/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	06/11/2022	06/11/2022
					CT3001	Implantable metal	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	06/11/2022	06/11/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	06/11/2022	06/11/2022
					CT3002	Implantable synthetic polymer	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	06/11/2022	06/11/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	06/11/2022	06/11/2022
					CT3006	Implantable inorganic material	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	06/11/2022	06/11/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	06/11/2022	06/11/2022
					CT314	Electrically powered	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	06/11/2022	06/11/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	06/11/2022	06/11/2022
					CT321	Long-term surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	06/11/2022	06/11/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	06/11/2022	06/11/2022
					CT449	Generators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	06/11/2022	06/11/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	06/11/2022	06/11/2022
					CT981	Single use	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	06/11/2022	06/11/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	06/11/2022	06/11/2022
					CT983	Surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	06/11/2022	06/11/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	06/11/2022	06/11/2022
11533	2022	All other reportable incidents	47265	Dual-chamber implantable pacemaker, rate-responsive	CT141	Electrodes/Leads and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	06/12/2022	06/12/2022
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	06/12/2022	06/12/2022
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	06/12/2022	06/12/2022
					CT145	Cardiovascular devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	06/12/2022	06/12/2022
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	06/12/2022	06/12/2022
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	06/12/2022	06/12/2022
					CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	06/12/2022	06/12/2022
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	06/12/2022	06/12/2022
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	06/12/2022	06/12/2022
					CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	06/12/2022	06/12/2022
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	06/12/2022	06/12/2022
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	06/12/2022	06/12/2022
					CT233	Surgical	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	06/12/2022	06/12/2022
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	06/12/2022	06/12/2022
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	06/12/2022	06/12/2022
					CT2406	Implantable	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	06/12/2022	06/12/2022
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	06/12/2022	06/12/2022
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	06/12/2022	06/12/2022
					CT2407	Active implantable	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	06/12/2022	06/12/2022
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	06/12/2022	06/12/2022
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	06/12/2022	06/12/2022
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	06/12/2022	06/12/2022
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	06/12/2022	06/12/2022

							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	06/12/2022	06/12/2022
					CT2727	Active implant devices/ancillaries	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	06/12/2022	06/12/2022
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	06/12/2022	06/12/2022
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	06/12/2022	06/12/2022
					CT2750	Implantable cardiac electrical stimulators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	06/12/2022	06/12/2022
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	06/12/2022	06/12/2022
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	06/12/2022	06/12/2022
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	06/12/2022	06/12/2022
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	06/12/2022	06/12/2022
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	06/12/2022	06/12/2022
					CT2985	Active	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	06/12/2022	06/12/2022
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	06/12/2022	06/12/2022
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	06/12/2022	06/12/2022
					CT3001	Implantable metal	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	06/12/2022	06/12/2022
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	06/12/2022	06/12/2022
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	06/12/2022	06/12/2022
					CT3002	Implantable synthetic polymer	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	06/12/2022	06/12/2022
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	06/12/2022	06/12/2022
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	06/12/2022	06/12/2022
					CT3006	Implantable inorganic material	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	06/12/2022	06/12/2022
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	06/12/2022	06/12/2022
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	06/12/2022	06/12/2022
					CT314	Electrically powered	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	06/12/2022	06/12/2022
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	06/12/2022	06/12/2022
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	06/12/2022	06/12/2022
					CT321	Long-term surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	06/12/2022	06/12/2022
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	06/12/2022	06/12/2022
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	06/12/2022	06/12/2022
					CT449	Generators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	06/12/2022	06/12/2022
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	06/12/2022	06/12/2022
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	06/12/2022	06/12/2022
					CT981	Single use	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	06/12/2022	06/12/2022
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	06/12/2022	06/12/2022
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	06/12/2022	06/12/2022
					CT983	Surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	06/12/2022	06/12/2022
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	06/12/2022	06/12/2022
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	06/12/2022	06/12/2022
11534	2022	All other reportable incidents			CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem		
11535	2022	All other reportable incidents	47265	Dual-chamber implantable pacemaker, rate-responsive	CT141	Electrodes/Leads and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	24/10/2022	24/10/2022
							F1901	Additional Surgery	A070504	Premature Discharge of Battery	24/10/2022	24/10/2022
					CT145	Cardiovascular devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	24/10/2022	24/10/2022
							F1901	Additional Surgery	A070504	Premature Discharge of Battery	24/10/2022	24/10/2022
					CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	24/10/2022	24/10/2022
							F1901	Additional Surgery	A070504	Premature Discharge of Battery	24/10/2022	24/10/2022
					CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	24/10/2022	24/10/2022
							F1901	Additional Surgery	A070504	Premature Discharge of Battery	24/10/2022	24/10/2022
					CT233	Surgical	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	24/10/2022	24/10/2022
							F1901	Additional Surgery	A070504	Premature Discharge of Battery	24/10/2022	24/10/2022
					CT2406	Implantable	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	24/10/2022	24/10/2022
							F1901	Additional Surgery	A070504	Premature Discharge of Battery	24/10/2022	24/10/2022
					CT2407	Active implantable	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	24/10/2022	24/10/2022
							F1901	Additional Surgery	A070504	Premature Discharge of Battery	24/10/2022	24/10/2022
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	24/10/2022	24/10/2022
							F1901	Additional Surgery	A070504	Premature Discharge of Battery	24/10/2022	24/10/2022
					CT2727	Active implant devices/ancillaries	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	24/10/2022	24/10/2022
							F1901	Additional Surgery	A070504	Premature Discharge of Battery	24/10/2022	24/10/2022
					CT2750	Implantable cardiac electrical stimulators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	24/10/2022	24/10/2022
							F1901	Additional Surgery	A070504	Premature Discharge of Battery	24/10/2022	24/10/2022
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	24/10/2022	24/10/2022
							F1901	Additional Surgery	A070504	Premature Discharge of Battery	24/10/2022	24/10/2022
					CT2985	Active	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	24/10/2022	24/10/2022
							F1901	Additional Surgery	A070504	Premature Discharge of Battery	24/10/2022	24/10/2022
					CT3001	Implantable metal	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	24/10/2022	24/10/2022
							F1901	Additional Surgery	A070504	Premature Discharge of Battery	24/10/2022	24/10/2022
					CT3002	Implantable synthetic polymer	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	24/10/2022	24/10/2022
							F1901	Additional Surgery	A070504	Premature Discharge of Battery	24/10/2022	24/10/2022
					CT3006	Implantable inorganic material	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	24/10/2022	24/10/2022
							F1901	Additional Surgery	A070504	Premature Discharge of Battery	24/10/2022	24/10/2022
					CT314	Electrically powered	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	24/10/2022	24/10/2022
							F1901	Additional Surgery	A070504	Premature Discharge of Battery	24/10/2022	24/10/2022
					CT321	Long-term surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	24/10/2022	24/10/2022
							F1901	Additional Surgery	A070504	Premature Discharge of Battery	24/10/2022	24/10/2022
					CT449	Generators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	24/10/2022	24/10/2022
							F1901	Additional Surgery	A070504	Premature Discharge of Battery	24/10/2022	24/10/2022
					CT981	Single use	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	24/10/2022	24/10/2022
							F1901	Additional Surgery	A070504	Premature Discharge of Battery	24/10/2022	24/10/2022
					CT983	Surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	24/10/2022	24/10/2022
							F1901	Additional Surgery	A070504	Premature Discharge of Battery	24/10/2022	24/10/2022
11536	2022	All other reportable incidents	47265	Dual-chamber implantable pacemaker, rate-responsive	CT141	Electrodes/Leads and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	18/10/2022	18/10/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	18/10/2022	18/10/2022
					CT145	Cardiovascular devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	18/10/2022	18/10/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	18/10/2022	18/10/2022
					CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	18/10/2022	18/10/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	18/10/2022	18/10/2022
					CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	18/10/2022	18/10/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	18/10/2022	18/10/2022
					CT233	Surgical	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	18/10/2022	18/10/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	18/10/2022	18/10/2022

					CT2406	Implantable	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	18/10/2022	18/10/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	18/10/2022	18/10/2022
					CT2407	Active implantable	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	18/10/2022	18/10/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	18/10/2022	18/10/2022
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	18/10/2022	18/10/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	18/10/2022	18/10/2022
					CT2727	Active implant devices/ancillaries	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	18/10/2022	18/10/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	18/10/2022	18/10/2022
					CT2750	Implantable cardiac electrical stimulators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	18/10/2022	18/10/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	18/10/2022	18/10/2022
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	18/10/2022	18/10/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	18/10/2022	18/10/2022
					CT2985	Active	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	18/10/2022	18/10/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	18/10/2022	18/10/2022
					CT3001	Implantable metal	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	18/10/2022	18/10/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	18/10/2022	18/10/2022
					CT3002	Implantable synthetic polymer	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	18/10/2022	18/10/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	18/10/2022	18/10/2022
					CT3006	Implantable inorganic material	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	18/10/2022	18/10/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	18/10/2022	18/10/2022
					CT314	Electrically powered	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	18/10/2022	18/10/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	18/10/2022	18/10/2022
					CT321	Long-term surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	18/10/2022	18/10/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	18/10/2022	18/10/2022
					CT449	Generators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	18/10/2022	18/10/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	18/10/2022	18/10/2022
					CT981	Single use	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	18/10/2022	18/10/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	18/10/2022	18/10/2022
					CT983	Surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	18/10/2022	18/10/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	18/10/2022	18/10/2022
11537	2022	All other reportable incidents	35852	Single-chamber implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E2104	Electric Shock	A070909	Oversensing		
									A071302	Inappropriate/Inadequate Shock/Stimulation		
									A090801	Signal Artifact/Noise		
							F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A070909	Oversensing		
									A071302	Inappropriate/Inadequate Shock/Stimulation		
									A090801	Signal Artifact/Noise		
11538	2022	All other reportable incidents	63188	Left-ventricular wireless pacing system sensor/transmitter	CT141	Electrodes/Leads and associated devices			A070801	Complete Loss of Power		
					CT145	Cardiovascular devices			A070801	Complete Loss of Power		
					CT1577	Wireless monitoring system transmitters			A070801	Complete Loss of Power		
					CT1684	Cardiac pacemakers and associated devices			A070801	Complete Loss of Power		
					CT2317	Transmitters/Receivers/Transceivers			A070801	Complete Loss of Power		
					CT233	Surgical			A070801	Complete Loss of Power		
					CT2406	Implantable			A070801	Complete Loss of Power		
					CT2407	Active implantable			A070801	Complete Loss of Power		
					CT255	Transducers and associated devices			A070801	Complete Loss of Power		
					CT2726	Physiological monitoring/measuring devices			A070801	Complete Loss of Power		
					CT2727	Active implant devices/ancillaries			A070801	Complete Loss of Power		
					CT2750	Implantable cardiac electrical stimulators and associated devices			A070801	Complete Loss of Power		
					CT2885	Physiological monitoring/measuring system transmitters/receivers/transceivers			A070801	Complete Loss of Power		
					CT2985	Active			A070801	Complete Loss of Power		
					CT3001	Implantable metal			A070801	Complete Loss of Power		
					CT3002	Implantable synthetic polymer			A070801	Complete Loss of Power		
					CT3006	Implantable inorganic material			A070801	Complete Loss of Power		
					CT314	Electrically powered			A070801	Complete Loss of Power		
					CT321	Long-term surgical invasive			A070801	Complete Loss of Power		
					CT384	Sensors and associated devices			A070801	Complete Loss of Power		
					CT449	Generators and associated devices			A070801	Complete Loss of Power		
					CT981	Single use			A070801	Complete Loss of Power		
					CT983	Surgical invasive			A070801	Complete Loss of Power		
					CT994	Sensors			A070801	Complete Loss of Power		
11539	2022	All other reportable incidents	47265	Dual-chamber implantable pacemaker, rate-responsive	CT141	Electrodes/Leads and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A090804	False Positive Result		
							F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A090804	False Positive Result		
					CT145	Cardiovascular devices	E2403	No Clinical Signs, Symptoms or Conditions	A090804	False Positive Result		
							F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A090804	False Positive Result		
					CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A090804	False Positive Result		
							F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A090804	False Positive Result		
					CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A090804	False Positive Result		
							F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A090804	False Positive Result		
					CT233	Surgical	E2403	No Clinical Signs, Symptoms or Conditions	A090804	False Positive Result		
							F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A090804	False Positive Result		
					CT2406	Implantable	E2403	No Clinical Signs, Symptoms or Conditions	A090804	False Positive Result		
							F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A090804	False Positive Result		
					CT2407	Active implantable	E2403	No Clinical Signs, Symptoms or Conditions	A090804	False Positive Result		
							F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A090804	False Positive Result		
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A090804	False Positive Result		
							F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A090804	False Positive Result		
					CT2727	Active implant devices/ancillaries	E2403	No Clinical Signs, Symptoms or Conditions	A090804	False Positive Result		
							F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A090804	False Positive Result		
					CT2750	Implantable cardiac electrical stimulators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A090804	False Positive Result		
							F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A090804	False Positive Result		
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A090804	False Positive Result		
							F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A090804	False Positive Result		
					CT2985	Active	E2403	No Clinical Signs, Symptoms or Conditions	A090804	False Positive Result		
							F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A090804	False Positive Result		
					CT3001	Implantable metal	E2403	No Clinical Signs, Symptoms or Conditions	A090804	False Positive Result		
							F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A090804	False Positive Result		

					CT3002	Implantable synthetic polymer	E2403	No Clinical Signs, Symptoms or Conditions	A090804	False Positive Result		
							F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A090804	False Positive Result		
					CT3006	Implantable inorganic material	E2403	No Clinical Signs, Symptoms or Conditions	A090804	False Positive Result		
							F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A090804	False Positive Result		
					CT314	Electrically powered	E2403	No Clinical Signs, Symptoms or Conditions	A090804	False Positive Result		
							F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A090804	False Positive Result		
					CT321	Long-term surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A090804	False Positive Result		
							F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A090804	False Positive Result		
					CT449	Generators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A090804	False Positive Result		
							F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A090804	False Positive Result		
					CT981	Single use	E2403	No Clinical Signs, Symptoms or Conditions	A090804	False Positive Result		
							F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A090804	False Positive Result		
					CT983	Surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A090804	False Positive Result		
							F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A090804	False Positive Result		
11540	2022	All other reportable incidents	35852	Single-chamber implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
11541	2022	All other reportable incidents	47265	Dual-chamber implantable pacemaker, rate-responsive	CT141	Electrodes/Leads and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
					CT145	Cardiovascular devices	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
					CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
					CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
					CT233	Surgical	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
					CT2406	Implantable	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
					CT2407	Active implantable	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
					CT2727	Active implant devices/ancillaries	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
					CT2750	Implantable cardiac electrical stimulators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
					CT2985	Active	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
					CT3001	Implantable metal	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
					CT3002	Implantable synthetic polymer	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
					CT3006	Implantable inorganic material	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
					CT314	Electrically powered	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
					CT321	Long-term surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
					CT449	Generators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		

					CT981	Single use	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
					CT983	Surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
11542	2022	All other reportable incidents	35852	Single-chamber implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem	24/11/2022	24/11/2022
									A070504	Premature Discharge of Battery	24/11/2022	24/11/2022
									A070909	Oversensing	24/11/2022	24/11/2022
									A071302	Inappropriate/Inadequate Shock/Stimulation	24/11/2022	24/11/2022
									A090801	Signal Artifact/Noise	24/11/2022	24/11/2022
									A1301	Failure to Read Input Signal	24/11/2022	24/11/2022
							F08	Hospitalization or Prolonged Hospitalization	A0705	Battery Problem	24/11/2022	24/11/2022
									A070504	Premature Discharge of Battery	24/11/2022	24/11/2022
									A070909	Oversensing	24/11/2022	24/11/2022
									A071302	Inappropriate/Inadequate Shock/Stimulation	24/11/2022	24/11/2022
									A090801	Signal Artifact/Noise	24/11/2022	24/11/2022
									A1301	Failure to Read Input Signal	24/11/2022	24/11/2022
							F1905	Device Revision or Replacement	A0705	Battery Problem	24/11/2022	24/11/2022
									A070504	Premature Discharge of Battery	24/11/2022	24/11/2022
									A070909	Oversensing	24/11/2022	24/11/2022
									A071302	Inappropriate/Inadequate Shock/Stimulation	24/11/2022	24/11/2022
									A090801	Signal Artifact/Noise	24/11/2022	24/11/2022
									A1301	Failure to Read Input Signal	24/11/2022	24/11/2022
11543	2022	All other reportable incidents			CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A0203	Defective Device		
							F08	Hospitalization or Prolonged Hospitalization	A0203	Defective Device		
11544	2022	All other reportable incidents	37265	Dual-chamber implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A072201	High Impedance		
									A072202	Low Impedance		
							F26	No Health Consequences or Impact	A072201	High Impedance		
									A072202	Low Impedance		
11545	2022	All other reportable incidents	35852	Single-chamber implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E2104	Electric Shock	A071302	Inappropriate/Inadequate Shock/Stimulation	04/08/2022	04/08/2022
							F08	Hospitalization or Prolonged Hospitalization	A071302	Inappropriate/Inadequate Shock/Stimulation	04/08/2022	04/08/2022
							F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A071302	Inappropriate/Inadequate Shock/Stimulation	04/08/2022	04/08/2022
							F1905	Device Revision or Replacement	A071302	Inappropriate/Inadequate Shock/Stimulation	04/08/2022	04/08/2022
11546	2022		47265	Dual-chamber implantable pacemaker, rate-responsive	CT141	Electrodes/Leads and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	06/11/2022	06/11/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	06/11/2022	06/11/2022
					CT145	Cardiovascular devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	06/11/2022	06/11/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	06/11/2022	06/11/2022
					CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	06/11/2022	06/11/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	06/11/2022	06/11/2022
					CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	06/11/2022	06/11/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	06/11/2022	06/11/2022
					CT233	Surgical	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	06/11/2022	06/11/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	06/11/2022	06/11/2022
					CT2406	Implantable	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	06/11/2022	06/11/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	06/11/2022	06/11/2022
					CT2407	Active implantable	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	06/11/2022	06/11/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	06/11/2022	06/11/2022
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	06/11/2022	06/11/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	06/11/2022	06/11/2022
					CT2727	Active implant devices/ancillaries	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	06/11/2022	06/11/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	06/11/2022	06/11/2022
					CT2750	Implantable cardiac electrical stimulators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	06/11/2022	06/11/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	06/11/2022	06/11/2022
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	06/11/2022	06/11/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	06/11/2022	06/11/2022
					CT2985	Active	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	06/11/2022	06/11/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	06/11/2022	06/11/2022
					CT3001	Implantable metal	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	06/11/2022	06/11/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	06/11/2022	06/11/2022
					CT3002	Implantable synthetic polymer	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	06/11/2022	06/11/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	06/11/2022	06/11/2022
					CT3006	Implantable inorganic material	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	06/11/2022	06/11/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	06/11/2022	06/11/2022
					CT314	Electrically powered	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	06/11/2022	06/11/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	06/11/2022	06/11/2022
					CT321	Long-term surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	06/11/2022	06/11/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	06/11/2022	06/11/2022
					CT449	Generators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	06/11/2022	06/11/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	06/11/2022	06/11/2022
					CT981	Single use	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	06/11/2022	06/11/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	06/11/2022	06/11/2022
					CT983	Surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	06/11/2022	06/11/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	06/11/2022	06/11/2022
11547	2022	All other reportable incidents	47265	Dual-chamber implantable pacemaker, rate-responsive	CT141	Electrodes/Leads and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	26/10/2022	26/10/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	26/10/2022	26/10/2022
					CT145	Cardiovascular devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	26/10/2022	26/10/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	26/10/2022	26/10/2022
					CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	26/10/2022	26/10/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	26/10/2022	26/10/2022
					CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	26/10/2022	26/10/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	26/10/2022	26/10/2022
					CT233	Surgical	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	26/10/2022	26/10/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	26/10/2022	26/10/2022
					CT2406	Implantable	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	26/10/2022	26/10/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	26/10/2022	26/10/2022

					CT2407	Active implantable	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	26/10/2022	26/10/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	26/10/2022	26/10/2022
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	26/10/2022	26/10/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	26/10/2022	26/10/2022
					CT2727	Active implant devices/ancillaries	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	26/10/2022	26/10/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	26/10/2022	26/10/2022
					CT2750	Implantable cardiac electrical stimulators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	26/10/2022	26/10/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	26/10/2022	26/10/2022
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	26/10/2022	26/10/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	26/10/2022	26/10/2022
					CT2985	Active	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	26/10/2022	26/10/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	26/10/2022	26/10/2022
					CT3001	Implantable metal	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	26/10/2022	26/10/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	26/10/2022	26/10/2022
					CT3002	Implantable synthetic polymer	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	26/10/2022	26/10/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	26/10/2022	26/10/2022
					CT3006	Implantable inorganic material	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	26/10/2022	26/10/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	26/10/2022	26/10/2022
					CT314	Electrically powered	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	26/10/2022	26/10/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	26/10/2022	26/10/2022
					CT321	Long-term surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	26/10/2022	26/10/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	26/10/2022	26/10/2022
					CT449	Generators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	26/10/2022	26/10/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	26/10/2022	26/10/2022
					CT981	Single use	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	26/10/2022	26/10/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	26/10/2022	26/10/2022
					CT983	Surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	26/10/2022	26/10/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	26/10/2022	26/10/2022
11548	2022		47265	Dual-chamber implantable pacemaker, rate-responsive	CT141	Electrodes/Leads and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	29/09/2022	29/09/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	29/09/2022	29/09/2022
					CT145	Cardiovascular devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	29/09/2022	29/09/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	29/09/2022	29/09/2022
					CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	29/09/2022	29/09/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	29/09/2022	29/09/2022
					CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	29/09/2022	29/09/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	29/09/2022	29/09/2022
					CT233	Surgical	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	29/09/2022	29/09/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	29/09/2022	29/09/2022
					CT2406	Implantable	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	29/09/2022	29/09/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	29/09/2022	29/09/2022
					CT2407	Active implantable	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	29/09/2022	29/09/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	29/09/2022	29/09/2022
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	29/09/2022	29/09/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	29/09/2022	29/09/2022
					CT2727	Active implant devices/ancillaries	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	29/09/2022	29/09/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	29/09/2022	29/09/2022
					CT2750	Implantable cardiac electrical stimulators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	29/09/2022	29/09/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	29/09/2022	29/09/2022
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	29/09/2022	29/09/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	29/09/2022	29/09/2022
					CT2985	Active	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	29/09/2022	29/09/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	29/09/2022	29/09/2022
					CT3001	Implantable metal	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	29/09/2022	29/09/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	29/09/2022	29/09/2022
					CT3002	Implantable synthetic polymer	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	29/09/2022	29/09/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	29/09/2022	29/09/2022
					CT3006	Implantable inorganic material	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	29/09/2022	29/09/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	29/09/2022	29/09/2022
					CT314	Electrically powered	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	29/09/2022	29/09/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	29/09/2022	29/09/2022
					CT321	Long-term surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	29/09/2022	29/09/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	29/09/2022	29/09/2022
					CT449	Generators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	29/09/2022	29/09/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	29/09/2022	29/09/2022
					CT981	Single use	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	29/09/2022	29/09/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	29/09/2022	29/09/2022
					CT983	Surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	29/09/2022	29/09/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	29/09/2022	29/09/2022
11549	2022	All other reportable incidents	47265	Dual-chamber implantable pacemaker, rate-responsive	CT141	Electrodes/Leads and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT145	Cardiovascular devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT233	Surgical	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT2406	Implantable	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT2407	Active implantable	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT2727	Active implant devices/ancillaries	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT2750	Implantable cardiac electrical stimulators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		

					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT2985	Active	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT3001	Implantable metal	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT3002	Implantable synthetic polymer	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT3006	Implantable inorganic material	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT314	Electrically powered	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT321	Long-term surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT449	Generators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT981	Single use	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT983	Surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
11550	2022		47265	Dual-chamber implantable pacemaker, rate-responsive	CT141	Electrodes/Leads and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	18/10/2022	18/10/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	18/10/2022	18/10/2022
					CT145	Cardiovascular devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	18/10/2022	18/10/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	18/10/2022	18/10/2022
					CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	18/10/2022	18/10/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	18/10/2022	18/10/2022
					CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	18/10/2022	18/10/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	18/10/2022	18/10/2022
					CT233	Surgical	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	18/10/2022	18/10/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	18/10/2022	18/10/2022
					CT2406	Implantable	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	18/10/2022	18/10/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	18/10/2022	18/10/2022
					CT2407	Active implantable	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	18/10/2022	18/10/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	18/10/2022	18/10/2022
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	18/10/2022	18/10/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	18/10/2022	18/10/2022
					CT2727	Active implant devices/ancillaries	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	18/10/2022	18/10/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	18/10/2022	18/10/2022
					CT2750	Implantable cardiac electrical stimulators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	18/10/2022	18/10/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	18/10/2022	18/10/2022
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	18/10/2022	18/10/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	18/10/2022	18/10/2022
					CT2985	Active	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	18/10/2022	18/10/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	18/10/2022	18/10/2022
					CT3001	Implantable metal	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	18/10/2022	18/10/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	18/10/2022	18/10/2022
					CT3002	Implantable synthetic polymer	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	18/10/2022	18/10/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	18/10/2022	18/10/2022
					CT3006	Implantable inorganic material	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	18/10/2022	18/10/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	18/10/2022	18/10/2022
					CT314	Electrically powered	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	18/10/2022	18/10/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	18/10/2022	18/10/2022
					CT321	Long-term surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	18/10/2022	18/10/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	18/10/2022	18/10/2022
					CT449	Generators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	18/10/2022	18/10/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	18/10/2022	18/10/2022
					CT981	Single use	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	18/10/2022	18/10/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	18/10/2022	18/10/2022
					CT983	Surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	18/10/2022	18/10/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	18/10/2022	18/10/2022
11551	2022		47265	Dual-chamber implantable pacemaker, rate-responsive	CT141	Electrodes/Leads and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	19/10/2022	19/10/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	19/10/2022	19/10/2022
					CT145	Cardiovascular devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	19/10/2022	19/10/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	19/10/2022	19/10/2022
					CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	19/10/2022	19/10/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	19/10/2022	19/10/2022
					CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	19/10/2022	19/10/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	19/10/2022	19/10/2022
					CT233	Surgical	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	19/10/2022	19/10/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	19/10/2022	19/10/2022
					CT2406	Implantable	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	19/10/2022	19/10/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	19/10/2022	19/10/2022
					CT2407	Active implantable	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	19/10/2022	19/10/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	19/10/2022	19/10/2022
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	19/10/2022	19/10/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	19/10/2022	19/10/2022
					CT2727	Active implant devices/ancillaries	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	19/10/2022	19/10/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	19/10/2022	19/10/2022
					CT2750	Implantable cardiac electrical stimulators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	19/10/2022	19/10/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	19/10/2022	19/10/2022
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	19/10/2022	19/10/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	19/10/2022	19/10/2022
					CT2985	Active	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	19/10/2022	19/10/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	19/10/2022	19/10/2022
					CT3001	Implantable metal	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	19/10/2022	19/10/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	19/10/2022	19/10/2022
					CT3002	Implantable synthetic polymer	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	19/10/2022	19/10/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	19/10/2022	19/10/2022

					CT3006	Implantable inorganic material	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	19/10/2022	19/10/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	19/10/2022	19/10/2022
					CT314	Electrically powered	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	19/10/2022	19/10/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	19/10/2022	19/10/2022
					CT321	Long-term surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	19/10/2022	19/10/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	19/10/2022	19/10/2022
					CT449	Generators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	19/10/2022	19/10/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	19/10/2022	19/10/2022
					CT981	Single use	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	19/10/2022	19/10/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	19/10/2022	19/10/2022
					CT983	Surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	19/10/2022	19/10/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	19/10/2022	19/10/2022
11552	2022	All other reportable incidents	35223	Endocardial/interventricular septal pacing lead	CT1684	Cardiac pacemakers and associated devices	E0112	Dizziness	A070909	Oversensing		
							F1901	Additional Surgery	A070909	Oversensing		
11553	2022		47265	Dual-chamber implantable pacemaker, rate-responsive	CT141	Electrodes/Leads and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	19/10/2022	19/10/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	19/10/2022	19/10/2022
					CT145	Cardiovascular devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	19/10/2022	19/10/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	19/10/2022	19/10/2022
					CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	19/10/2022	19/10/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	19/10/2022	19/10/2022
					CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	19/10/2022	19/10/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	19/10/2022	19/10/2022
					CT233	Surgical	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	19/10/2022	19/10/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	19/10/2022	19/10/2022
					CT2406	Implantable	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	19/10/2022	19/10/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	19/10/2022	19/10/2022
					CT2407	Active implantable	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	19/10/2022	19/10/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	19/10/2022	19/10/2022
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	19/10/2022	19/10/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	19/10/2022	19/10/2022
					CT2727	Active implant devices/ancillaries	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	19/10/2022	19/10/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	19/10/2022	19/10/2022
					CT2750	Implantable cardiac electrical stimulators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	19/10/2022	19/10/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	19/10/2022	19/10/2022
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	19/10/2022	19/10/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	19/10/2022	19/10/2022
					CT2985	Active	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	19/10/2022	19/10/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	19/10/2022	19/10/2022
					CT3001	Implantable metal	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	19/10/2022	19/10/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	19/10/2022	19/10/2022
					CT3002	Implantable synthetic polymer	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	19/10/2022	19/10/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	19/10/2022	19/10/2022
					CT3006	Implantable inorganic material	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	19/10/2022	19/10/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	19/10/2022	19/10/2022
					CT314	Electrically powered	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	19/10/2022	19/10/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	19/10/2022	19/10/2022
					CT321	Long-term surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	19/10/2022	19/10/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	19/10/2022	19/10/2022
					CT449	Generators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	19/10/2022	19/10/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	19/10/2022	19/10/2022
					CT981	Single use	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	19/10/2022	19/10/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	19/10/2022	19/10/2022
					CT983	Surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	19/10/2022	19/10/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	19/10/2022	19/10/2022
11554	2022	All other reportable incidents	47265	Dual-chamber implantable pacemaker, rate-responsive	CT141	Electrodes/Leads and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	19/10/2022	19/10/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	19/10/2022	19/10/2022
					CT145	Cardiovascular devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	19/10/2022	19/10/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	19/10/2022	19/10/2022
					CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	19/10/2022	19/10/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	19/10/2022	19/10/2022
					CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	19/10/2022	19/10/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	19/10/2022	19/10/2022
					CT233	Surgical	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	19/10/2022	19/10/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	19/10/2022	19/10/2022
					CT2406	Implantable	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	19/10/2022	19/10/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	19/10/2022	19/10/2022
					CT2407	Active implantable	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	19/10/2022	19/10/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	19/10/2022	19/10/2022
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	19/10/2022	19/10/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	19/10/2022	19/10/2022
					CT2727	Active implant devices/ancillaries	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	19/10/2022	19/10/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	19/10/2022	19/10/2022
					CT2750	Implantable cardiac electrical stimulators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	19/10/2022	19/10/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	19/10/2022	19/10/2022
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	19/10/2022	19/10/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	19/10/2022	19/10/2022
					CT2985	Active	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	19/10/2022	19/10/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	19/10/2022	19/10/2022
					CT3001	Implantable metal	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	19/10/2022	19/10/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	19/10/2022	19/10/2022
					CT3002	Implantable synthetic polymer	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	19/10/2022	19/10/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	19/10/2022	19/10/2022
					CT3006	Implantable inorganic material	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	19/10/2022	19/10/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	19/10/2022	19/10/2022
					CT314	Electrically powered	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	19/10/2022	19/10/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	19/10/2022	19/10/2022
					CT321	Long-term surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	19/10/2022	19/10/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	19/10/2022	19/10/2022

					CT449	Generators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	19/10/2022	19/10/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	19/10/2022	19/10/2022
					CT981	Single use	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	19/10/2022	19/10/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	19/10/2022	19/10/2022
					CT983	Surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	19/10/2022	19/10/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	19/10/2022	19/10/2022
11555	2022	All other reportable incidents	17882	Physiologic-monitoring defibrillation system	CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A07	Electrical /Electronic Property Problem		
									A0721	Circuit Failure		
							F27	Problem Identified During Non-Clinical Procedure	A07	Electrical /Electronic Property Problem		
									A0721	Circuit Failure		
11556	2022	All other reportable incidents	47265	Dual-chamber implantable pacemaker, rate-responsive	CT141	Electrodes/Leads and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	19/10/2022	19/10/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	19/10/2022	19/10/2022
					CT145	Cardiovascular devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	19/10/2022	19/10/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	19/10/2022	19/10/2022
					CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	19/10/2022	19/10/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	19/10/2022	19/10/2022
					CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	19/10/2022	19/10/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	19/10/2022	19/10/2022
					CT233	Surgical	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	19/10/2022	19/10/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	19/10/2022	19/10/2022
					CT2406	Implantable	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	19/10/2022	19/10/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	19/10/2022	19/10/2022
					CT2407	Active implantable	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	19/10/2022	19/10/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	19/10/2022	19/10/2022
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	19/10/2022	19/10/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	19/10/2022	19/10/2022
					CT2727	Active implant devices/ancillaries	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	19/10/2022	19/10/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	19/10/2022	19/10/2022
					CT2750	Implantable cardiac electrical stimulators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	19/10/2022	19/10/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	19/10/2022	19/10/2022
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	19/10/2022	19/10/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	19/10/2022	19/10/2022
					CT2985	Active	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	19/10/2022	19/10/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	19/10/2022	19/10/2022
					CT3001	Implantable metal	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	19/10/2022	19/10/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	19/10/2022	19/10/2022
					CT3002	Implantable synthetic polymer	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	19/10/2022	19/10/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	19/10/2022	19/10/2022
					CT3006	Implantable inorganic material	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	19/10/2022	19/10/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	19/10/2022	19/10/2022
					CT314	Electrically powered	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	19/10/2022	19/10/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	19/10/2022	19/10/2022
					CT321	Long-term surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	19/10/2022	19/10/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	19/10/2022	19/10/2022
					CT449	Generators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	19/10/2022	19/10/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	19/10/2022	19/10/2022
					CT981	Single use	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	19/10/2022	19/10/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	19/10/2022	19/10/2022
					CT983	Surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	19/10/2022	19/10/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	19/10/2022	19/10/2022
11557	2022	All other reportable incidents	35852	Single-chamber implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E2104	Electric Shock	A070909	Oversensing		
									A071201	Failure to Convert Rhythm		
									A071302	Inappropriate/Inadequate Shock/Stimulation		
									A090801	Signal Artifact/Noise		
									A1301	Failure to Read Input Signal		
							F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A070909	Oversensing		
									A071201	Failure to Convert Rhythm		
									A071302	Inappropriate/Inadequate Shock/Stimulation		
									A090801	Signal Artifact/Noise		
									A1301	Failure to Read Input Signal		
11558	2022	All other reportable incidents	35852	Single-chamber implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E2104	Electric Shock	A070909	Oversensing		
									A071302	Inappropriate/Inadequate Shock/Stimulation		
									A1301	Failure to Read Input Signal		
							F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A070909	Oversensing		
									A071302	Inappropriate/Inadequate Shock/Stimulation		
									A1301	Failure to Read Input Signal		
11559	2022	All other reportable incidents	36241	Pacemaker/defibrillator lead	CT1684	Cardiac pacemakers and associated devices	E2104	Electric Shock	A070909	Oversensing		
									A071302	Inappropriate/Inadequate Shock/Stimulation		
									A1301	Failure to Read Input Signal		
							F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A070909	Oversensing		
									A071302	Inappropriate/Inadequate Shock/Stimulation		
									A1301	Failure to Read Input Signal		
11560	2022	All other reportable incidents	36241	Pacemaker/defibrillator lead	CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A071302	Inappropriate/Inadequate Shock/Stimulation		
									A072201	High Impedance		
									A072202	Low Impedance		
							F26	No Health Consequences or Impact	A071302	Inappropriate/Inadequate Shock/Stimulation		
									A072201	High Impedance		
									A072202	Low Impedance		
11561	2022	All other reportable incidents	36241	Pacemaker/defibrillator lead	CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A0401	Break		
									A040101	Fracture		
									A0709	Device Sensing Problem		
									A070909	Oversensing		
									A0908	Incorrect, Inadequate or Imprecise Result or Readings		
									A090801	Signal Artifact/Noise		
							F26	No Health Consequences or Impact	A0401	Break		
									A040101	Fracture		
									A0709	Device Sensing Problem		

									A070909	Oversensing		
									A0908	Incorrect, Inadequate or Imprecise Result or Readings		
									A090801	Signal Artifact/Noise		
11562	2022	All other reportable incidents	47265	Dual-chamber implantable pacemaker, rate-responsive	CT141	Electrodes/Leads and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A0908	Incorrect, Inadequate or Imprecise Result or Readings	11/10/2022	11/10/2022
							F08	Hospitalization or Prolonged Hospitalization	A0908	Incorrect, Inadequate or Imprecise Result or Readings	11/10/2022	11/10/2022
							F1905	Device Revision or Replacement	A0908	Incorrect, Inadequate or Imprecise Result or Readings	11/10/2022	11/10/2022
					CT145	Cardiovascular devices	E2403	No Clinical Signs, Symptoms or Conditions	A0908	Incorrect, Inadequate or Imprecise Result or Readings	11/10/2022	11/10/2022
							F08	Hospitalization or Prolonged Hospitalization	A0908	Incorrect, Inadequate or Imprecise Result or Readings	11/10/2022	11/10/2022
							F1905	Device Revision or Replacement	A0908	Incorrect, Inadequate or Imprecise Result or Readings	11/10/2022	11/10/2022
					CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A0908	Incorrect, Inadequate or Imprecise Result or Readings	11/10/2022	11/10/2022
							F08	Hospitalization or Prolonged Hospitalization	A0908	Incorrect, Inadequate or Imprecise Result or Readings	11/10/2022	11/10/2022
							F1905	Device Revision or Replacement	A0908	Incorrect, Inadequate or Imprecise Result or Readings	11/10/2022	11/10/2022
					CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A0908	Incorrect, Inadequate or Imprecise Result or Readings	11/10/2022	11/10/2022
							F08	Hospitalization or Prolonged Hospitalization	A0908	Incorrect, Inadequate or Imprecise Result or Readings	11/10/2022	11/10/2022
							F1905	Device Revision or Replacement	A0908	Incorrect, Inadequate or Imprecise Result or Readings	11/10/2022	11/10/2022
					CT233	Surgical	E2403	No Clinical Signs, Symptoms or Conditions	A0908	Incorrect, Inadequate or Imprecise Result or Readings	11/10/2022	11/10/2022
							F08	Hospitalization or Prolonged Hospitalization	A0908	Incorrect, Inadequate or Imprecise Result or Readings	11/10/2022	11/10/2022
							F1905	Device Revision or Replacement	A0908	Incorrect, Inadequate or Imprecise Result or Readings	11/10/2022	11/10/2022
					CT2406	Implantable	E2403	No Clinical Signs, Symptoms or Conditions	A0908	Incorrect, Inadequate or Imprecise Result or Readings	11/10/2022	11/10/2022
							F08	Hospitalization or Prolonged Hospitalization	A0908	Incorrect, Inadequate or Imprecise Result or Readings	11/10/2022	11/10/2022
							F1905	Device Revision or Replacement	A0908	Incorrect, Inadequate or Imprecise Result or Readings	11/10/2022	11/10/2022
					CT2407	Active implantable	E2403	No Clinical Signs, Symptoms or Conditions	A0908	Incorrect, Inadequate or Imprecise Result or Readings	11/10/2022	11/10/2022
							F08	Hospitalization or Prolonged Hospitalization	A0908	Incorrect, Inadequate or Imprecise Result or Readings	11/10/2022	11/10/2022
							F1905	Device Revision or Replacement	A0908	Incorrect, Inadequate or Imprecise Result or Readings	11/10/2022	11/10/2022
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A0908	Incorrect, Inadequate or Imprecise Result or Readings	11/10/2022	11/10/2022
							F08	Hospitalization or Prolonged Hospitalization	A0908	Incorrect, Inadequate or Imprecise Result or Readings	11/10/2022	11/10/2022
							F1905	Device Revision or Replacement	A0908	Incorrect, Inadequate or Imprecise Result or Readings	11/10/2022	11/10/2022
					CT2727	Active implant devices/ancillaries	E2403	No Clinical Signs, Symptoms or Conditions	A0908	Incorrect, Inadequate or Imprecise Result or Readings	11/10/2022	11/10/2022
							F08	Hospitalization or Prolonged Hospitalization	A0908	Incorrect, Inadequate or Imprecise Result or Readings	11/10/2022	11/10/2022
							F1905	Device Revision or Replacement	A0908	Incorrect, Inadequate or Imprecise Result or Readings	11/10/2022	11/10/2022
					CT2750	Implantable cardiac electrical stimulators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A0908	Incorrect, Inadequate or Imprecise Result or Readings	11/10/2022	11/10/2022
							F08	Hospitalization or Prolonged Hospitalization	A0908	Incorrect, Inadequate or Imprecise Result or Readings	11/10/2022	11/10/2022
							F1905	Device Revision or Replacement	A0908	Incorrect, Inadequate or Imprecise Result or Readings	11/10/2022	11/10/2022
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A0908	Incorrect, Inadequate or Imprecise Result or Readings	11/10/2022	11/10/2022
							F08	Hospitalization or Prolonged Hospitalization	A0908	Incorrect, Inadequate or Imprecise Result or Readings	11/10/2022	11/10/2022
							F1905	Device Revision or Replacement	A0908	Incorrect, Inadequate or Imprecise Result or Readings	11/10/2022	11/10/2022
					CT2985	Active	E2403	No Clinical Signs, Symptoms or Conditions	A0908	Incorrect, Inadequate or Imprecise Result or Readings	11/10/2022	11/10/2022
							F08	Hospitalization or Prolonged Hospitalization	A0908	Incorrect, Inadequate or Imprecise Result or Readings	11/10/2022	11/10/2022
							F1905	Device Revision or Replacement	A0908	Incorrect, Inadequate or Imprecise Result or Readings	11/10/2022	11/10/2022
					CT3001	Implantable metal	E2403	No Clinical Signs, Symptoms or Conditions	A0908	Incorrect, Inadequate or Imprecise Result or Readings	11/10/2022	11/10/2022
							F08	Hospitalization or Prolonged Hospitalization	A0908	Incorrect, Inadequate or Imprecise Result or Readings	11/10/2022	11/10/2022
							F1905	Device Revision or Replacement	A0908	Incorrect, Inadequate or Imprecise Result or Readings	11/10/2022	11/10/2022
					CT3002	Implantable synthetic polymer	E2403	No Clinical Signs, Symptoms or Conditions	A0908	Incorrect, Inadequate or Imprecise Result or Readings	11/10/2022	11/10/2022
							F08	Hospitalization or Prolonged Hospitalization	A0908	Incorrect, Inadequate or Imprecise Result or Readings	11/10/2022	11/10/2022
							F1905	Device Revision or Replacement	A0908	Incorrect, Inadequate or Imprecise Result or Readings	11/10/2022	11/10/2022
					CT3006	Implantable inorganic material	E2403	No Clinical Signs, Symptoms or Conditions	A0908	Incorrect, Inadequate or Imprecise Result or Readings	11/10/2022	11/10/2022
							F08	Hospitalization or Prolonged Hospitalization	A0908	Incorrect, Inadequate or Imprecise Result or Readings	11/10/2022	11/10/2022
							F1905	Device Revision or Replacement	A0908	Incorrect, Inadequate or Imprecise Result or Readings	11/10/2022	11/10/2022
					CT314	Electrically powered	E2403	No Clinical Signs, Symptoms or Conditions	A0908	Incorrect, Inadequate or Imprecise Result or Readings	11/10/2022	11/10/2022
							F08	Hospitalization or Prolonged Hospitalization	A0908	Incorrect, Inadequate or Imprecise Result or Readings	11/10/2022	11/10/2022

						F1905	Device Revision or Replacement	A0908	Incorrect, Inadequate or Imprecise Result or Readings	11/10/2022	11/10/2022
					CT321	Long-term surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A0908	Incorrect, Inadequate or Imprecise Result or Readings	11/10/2022	11/10/2022
							F08	Hospitalization or Prolonged Hospitalization	A0908	Incorrect, Inadequate or Imprecise Result or Readings	11/10/2022	11/10/2022
							F1905	Device Revision or Replacement	A0908	Incorrect, Inadequate or Imprecise Result or Readings	11/10/2022	11/10/2022
					CT449	Generators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A0908	Incorrect, Inadequate or Imprecise Result or Readings	11/10/2022	11/10/2022
							F08	Hospitalization or Prolonged Hospitalization	A0908	Incorrect, Inadequate or Imprecise Result or Readings	11/10/2022	11/10/2022
							F1905	Device Revision or Replacement	A0908	Incorrect, Inadequate or Imprecise Result or Readings	11/10/2022	11/10/2022
					CT981	Single use	E2403	No Clinical Signs, Symptoms or Conditions	A0908	Incorrect, Inadequate or Imprecise Result or Readings	11/10/2022	11/10/2022
							F08	Hospitalization or Prolonged Hospitalization	A0908	Incorrect, Inadequate or Imprecise Result or Readings	11/10/2022	11/10/2022
							F1905	Device Revision or Replacement	A0908	Incorrect, Inadequate or Imprecise Result or Readings	11/10/2022	11/10/2022
					CT983	Surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A0908	Incorrect, Inadequate or Imprecise Result or Readings	11/10/2022	11/10/2022
							F08	Hospitalization or Prolonged Hospitalization	A0908	Incorrect, Inadequate or Imprecise Result or Readings	11/10/2022	11/10/2022
							F1905	Device Revision or Replacement	A0908	Incorrect, Inadequate or Imprecise Result or Readings	11/10/2022	11/10/2022
11563	2022	All other reportable incidents	47265	Dual-chamber implantable pacemaker, rate-responsive	CT141	Electrodes/Leads and associated devices	E011903	Syncope/Fainting	A070101	Failure to Capture		
									A070102	High Capture Threshold		
									A12	Connection Problem		
							E060104	Bradycardia	A070101	Failure to Capture		
									A070102	High Capture Threshold		
									A12	Connection Problem		
							F08	Hospitalization or Prolonged Hospitalization	A070101	Failure to Capture		
									A070102	High Capture Threshold		
									A12	Connection Problem		
							F1904	Device Repositioning	A070101	Failure to Capture		
									A070102	High Capture Threshold		
									A12	Connection Problem		
					CT145	Cardiovascular devices	E011903	Syncope/Fainting	A070101	Failure to Capture		
									A070102	High Capture Threshold		
									A12	Connection Problem		
							E060104	Bradycardia	A070101	Failure to Capture		
									A070102	High Capture Threshold		
									A12	Connection Problem		
							F08	Hospitalization or Prolonged Hospitalization	A070101	Failure to Capture		
									A070102	High Capture Threshold		
									A12	Connection Problem		
							F1904	Device Repositioning	A070101	Failure to Capture		
									A070102	High Capture Threshold		
									A12	Connection Problem		
					CT1684	Cardiac pacemakers and associated devices	E011903	Syncope/Fainting	A070101	Failure to Capture		
									A070102	High Capture Threshold		
									A12	Connection Problem		
							E060104	Bradycardia	A070101	Failure to Capture		
									A070102	High Capture Threshold		
									A12	Connection Problem		
							F08	Hospitalization or Prolonged Hospitalization	A070101	Failure to Capture		
									A070102	High Capture Threshold		
									A12	Connection Problem		
							F1904	Device Repositioning	A070101	Failure to Capture		
									A070102	High Capture Threshold		
									A12	Connection Problem		
					CT1685	Cardiac pacemakers	E011903	Syncope/Fainting	A070101	Failure to Capture		
									A070102	High Capture Threshold		
									A12	Connection Problem		
							E060104	Bradycardia	A070101	Failure to Capture		
									A070102	High Capture Threshold		
									A12	Connection Problem		
							F08	Hospitalization or Prolonged Hospitalization	A070101	Failure to Capture		
									A070102	High Capture Threshold		
									A12	Connection Problem		
							F1904	Device Repositioning	A070101	Failure to Capture		
									A070102	High Capture Threshold		
									A12	Connection Problem		
					CT233	Surgical	E011903	Syncope/Fainting	A070101	Failure to Capture		
									A070102	High Capture Threshold		
									A12	Connection Problem		
							E060104	Bradycardia	A070101	Failure to Capture		
									A070102	High Capture Threshold		
									A12	Connection Problem		
							F08	Hospitalization or Prolonged Hospitalization	A070101	Failure to Capture		
									A070102	High Capture Threshold		
									A12	Connection Problem		
							F1904	Device Repositioning	A070101	Failure to Capture		
									A070102	High Capture Threshold		
									A12	Connection Problem		
					CT2406	Implantable	E011903	Syncope/Fainting	A070101	Failure to Capture		
									A070102	High Capture Threshold		
									A12	Connection Problem		
							E060104	Bradycardia	A070101	Failure to Capture		
									A070102	High Capture Threshold		

									A12	Connection Problem		
							F08	Hospitalization or Prolonged Hospitalization	A070101	Failure to Capture		
									A070102	High Capture Threshold		
									A12	Connection Problem		
							F1904	Device Repositioning	A070101	Failure to Capture		
									A070102	High Capture Threshold		
									A12	Connection Problem		
					CT2407	Active implantable	E011903	Syncope/Fainting	A070101	Failure to Capture		
									A070102	High Capture Threshold		
									A12	Connection Problem		
							E060104	Bradycardia	A070101	Failure to Capture		
									A070102	High Capture Threshold		
									A12	Connection Problem		
							F08	Hospitalization or Prolonged Hospitalization	A070101	Failure to Capture		
									A070102	High Capture Threshold		
									A12	Connection Problem		
							F1904	Device Repositioning	A070101	Failure to Capture		
									A070102	High Capture Threshold		
									A12	Connection Problem		
					CT257	Implantable cardiac pacemakers	E011903	Syncope/Fainting	A070101	Failure to Capture		
									A070102	High Capture Threshold		
									A12	Connection Problem		
							E060104	Bradycardia	A070101	Failure to Capture		
									A070102	High Capture Threshold		
									A12	Connection Problem		
							F08	Hospitalization or Prolonged Hospitalization	A070101	Failure to Capture		
									A070102	High Capture Threshold		
									A12	Connection Problem		
							F1904	Device Repositioning	A070101	Failure to Capture		
									A070102	High Capture Threshold		
									A12	Connection Problem		
					CT2727	Active implant devices/ancillaries	E011903	Syncope/Fainting	A070101	Failure to Capture		
									A070102	High Capture Threshold		
									A12	Connection Problem		
							E060104	Bradycardia	A070101	Failure to Capture		
									A070102	High Capture Threshold		
									A12	Connection Problem		
							F08	Hospitalization or Prolonged Hospitalization	A070101	Failure to Capture		
									A070102	High Capture Threshold		
									A12	Connection Problem		
							F1904	Device Repositioning	A070101	Failure to Capture		
									A070102	High Capture Threshold		
									A12	Connection Problem		
					CT2750	Implantable cardiac electrical stimulators and associated devices	E011903	Syncope/Fainting	A070101	Failure to Capture		
									A070102	High Capture Threshold		
									A12	Connection Problem		
							E060104	Bradycardia	A070101	Failure to Capture		
									A070102	High Capture Threshold		
									A12	Connection Problem		
							F08	Hospitalization or Prolonged Hospitalization	A070101	Failure to Capture		
									A070102	High Capture Threshold		
									A12	Connection Problem		
							F1904	Device Repositioning	A070101	Failure to Capture		
									A070102	High Capture Threshold		
									A12	Connection Problem		
					CT2866	Implantable cardiac electrical stimulators	E011903	Syncope/Fainting	A070101	Failure to Capture		
									A070102	High Capture Threshold		
									A12	Connection Problem		
							E060104	Bradycardia	A070101	Failure to Capture		
									A070102	High Capture Threshold		
									A12	Connection Problem		
							F08	Hospitalization or Prolonged Hospitalization	A070101	Failure to Capture		
									A070102	High Capture Threshold		
									A12	Connection Problem		
							F1904	Device Repositioning	A070101	Failure to Capture		
									A070102	High Capture Threshold		
									A12	Connection Problem		
					CT2985	Active	E011903	Syncope/Fainting	A070101	Failure to Capture		
									A070102	High Capture Threshold		
									A12	Connection Problem		
							E060104	Bradycardia	A070101	Failure to Capture		
									A070102	High Capture Threshold		
									A12	Connection Problem		
							F08	Hospitalization or Prolonged Hospitalization	A070101	Failure to Capture		
									A070102	High Capture Threshold		
									A12	Connection Problem		
							F1904	Device Repositioning	A070101	Failure to Capture		
									A070102	High Capture Threshold		
									A12	Connection Problem		
					CT3001	Implantable metal	E011903	Syncope/Fainting	A070101	Failure to Capture		
									A070102	High Capture Threshold		
									A12	Connection Problem		
							E060104	Bradycardia	A070101	Failure to Capture		
									A070102	High Capture Threshold		
									A12	Connection Problem		
							F08	Hospitalization or Prolonged Hospitalization	A070101	Failure to Capture		
									A070102	High Capture Threshold		
									A12	Connection Problem		

							F1904	Device Repositioning	A070101	Failure to Capture		
									A070102	High Capture Threshold		
									A12	Connection Problem		
					CT3002	Implantable synthetic polymer	E011903	Syncope/Fainting	A070101	Failure to Capture		
									A070102	High Capture Threshold		
									A12	Connection Problem		
							E060104	Bradycardia	A070101	Failure to Capture		
									A070102	High Capture Threshold		
									A12	Connection Problem		
							F08	Hospitalization or Prolonged Hospitalization	A070101	Failure to Capture		
									A070102	High Capture Threshold		
									A12	Connection Problem		
							F1904	Device Repositioning	A070101	Failure to Capture		
									A070102	High Capture Threshold		
									A12	Connection Problem		
					CT3006	Implantable inorganic material	E011903	Syncope/Fainting	A070101	Failure to Capture		
									A070102	High Capture Threshold		
									A12	Connection Problem		
							E060104	Bradycardia	A070101	Failure to Capture		
									A070102	High Capture Threshold		
									A12	Connection Problem		
							F08	Hospitalization or Prolonged Hospitalization	A070101	Failure to Capture		
									A070102	High Capture Threshold		
									A12	Connection Problem		
							F1904	Device Repositioning	A070101	Failure to Capture		
									A070102	High Capture Threshold		
									A12	Connection Problem		
					CT314	Electrically powered	E011903	Syncope/Fainting	A070101	Failure to Capture		
									A070102	High Capture Threshold		
									A12	Connection Problem		
							E060104	Bradycardia	A070101	Failure to Capture		
									A070102	High Capture Threshold		
									A12	Connection Problem		
							F08	Hospitalization or Prolonged Hospitalization	A070101	Failure to Capture		
									A070102	High Capture Threshold		
									A12	Connection Problem		
							F1904	Device Repositioning	A070101	Failure to Capture		
									A070102	High Capture Threshold		
									A12	Connection Problem		
					CT321	Long-term surgical invasive	E011903	Syncope/Fainting	A070101	Failure to Capture		
									A070102	High Capture Threshold		
									A12	Connection Problem		
							E060104	Bradycardia	A070101	Failure to Capture		
									A070102	High Capture Threshold		
									A12	Connection Problem		
							F08	Hospitalization or Prolonged Hospitalization	A070101	Failure to Capture		
									A070102	High Capture Threshold		
									A12	Connection Problem		
							F1904	Device Repositioning	A070101	Failure to Capture		
									A070102	High Capture Threshold		
									A12	Connection Problem		
					CT449	Generators and associated devices	E011903	Syncope/Fainting	A070101	Failure to Capture		
									A070102	High Capture Threshold		
									A12	Connection Problem		
							E060104	Bradycardia	A070101	Failure to Capture		
									A070102	High Capture Threshold		
									A12	Connection Problem		
							F08	Hospitalization or Prolonged Hospitalization	A070101	Failure to Capture		
									A070102	High Capture Threshold		
									A12	Connection Problem		
							F1904	Device Repositioning	A070101	Failure to Capture		
									A070102	High Capture Threshold		
									A12	Connection Problem		
					CT981	Single use	E011903	Syncope/Fainting	A070101	Failure to Capture		
									A070102	High Capture Threshold		
									A12	Connection Problem		
							E060104	Bradycardia	A070101	Failure to Capture		
									A070102	High Capture Threshold		
									A12	Connection Problem		
							F08	Hospitalization or Prolonged Hospitalization	A070101	Failure to Capture		
									A070102	High Capture Threshold		
									A12	Connection Problem		
							F1904	Device Repositioning	A070101	Failure to Capture		
									A070102	High Capture Threshold		
									A12	Connection Problem		
					CT983	Surgical invasive	E011903	Syncope/Fainting	A070101	Failure to Capture		
									A070102	High Capture Threshold		
									A12	Connection Problem		
							E060104	Bradycardia	A070101	Failure to Capture		
									A070102	High Capture Threshold		
									A12	Connection Problem		
							F08	Hospitalization or Prolonged Hospitalization	A070101	Failure to Capture		
									A070102	High Capture Threshold		
									A12	Connection Problem		
							F1904	Device Repositioning	A070101	Failure to Capture		
									A070102	High Capture Threshold		
									A12	Connection Problem		
11564	2022	All other reportable incidents	47265	Dual-chamber implantable pacemaker, rate-	CT141	Electrodes/Leads and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		

				responsive			F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT145	Cardiovascular devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT233	Surgical	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT2406	Implantable	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT2407	Active implantable	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT2727	Active implant devices/ancillaries	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT2750	Implantable cardiac electrical stimulators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT2985	Active	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT3001	Implantable metal	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT3002	Implantable synthetic polymer	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT3006	Implantable inorganic material	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT314	Electrically powered	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT321	Long-term surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT449	Generators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT981	Single use	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT983	Surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
11565	2022	All other reportable incidents	17882	Physiologic-monitoring defibrillation system	CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A150105	Key or Button Unresponsive/not Working		
							F27	Problem Identified During Non-Clinical Procedure	A150105	Key or Button Unresponsive/not Working		
11566	2022	All other reportable incidents	17882	Physiologic-monitoring defibrillation system	CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A071301	Failure to Deliver Shock/Stimulation		
							F26	No Health Consequences or Impact	A071301	Failure to Deliver Shock/Stimulation		
11567	2022	All other reportable incidents	35852	Single-chamber implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
11568	2022	All other reportable incidents	47263	Cardiac resynchronization therapy implantable pacemaker	CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
11569	2022	All other reportable incidents	47265	Dual-chamber implantable pacemaker, rate-responsive	CT141	Electrodes/Leads and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A071102	Failure to Interrogate	24/10/2022	24/10/2022
									A071204	No Pacing	24/10/2022	24/10/2022
							F1901	Additional Surgery	A071102	Failure to Interrogate	24/10/2022	24/10/2022
									A071204	No Pacing	24/10/2022	24/10/2022
					CT145	Cardiovascular devices	E2403	No Clinical Signs, Symptoms or Conditions	A071102	Failure to Interrogate	24/10/2022	24/10/2022
									A071204	No Pacing	24/10/2022	24/10/2022
							F1901	Additional Surgery	A071102	Failure to Interrogate	24/10/2022	24/10/2022
									A071204	No Pacing	24/10/2022	24/10/2022
					CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A071102	Failure to Interrogate	24/10/2022	24/10/2022
									A071204	No Pacing	24/10/2022	24/10/2022
							F1901	Additional Surgery	A071102	Failure to Interrogate	24/10/2022	24/10/2022
									A071204	No Pacing	24/10/2022	24/10/2022
					CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A071102	Failure to Interrogate	24/10/2022	24/10/2022
									A071204	No Pacing	24/10/2022	24/10/2022
							F1901	Additional Surgery	A071102	Failure to Interrogate	24/10/2022	24/10/2022
									A071204	No Pacing	24/10/2022	24/10/2022
					CT233	Surgical	E2403	No Clinical Signs, Symptoms or Conditions	A071102	Failure to Interrogate	24/10/2022	24/10/2022
									A071204	No Pacing	24/10/2022	24/10/2022
							F1901	Additional Surgery	A071102	Failure to Interrogate	24/10/2022	24/10/2022
									A071204	No Pacing	24/10/2022	24/10/2022
					CT2406	Implantable	E2403	No Clinical Signs, Symptoms or Conditions	A071102	Failure to Interrogate	24/10/2022	24/10/2022
									A071204	No Pacing	24/10/2022	24/10/2022
							F1901	Additional Surgery	A071102	Failure to Interrogate	24/10/2022	24/10/2022
									A071204	No Pacing	24/10/2022	24/10/2022
					CT2407	Active implantable	E2403	No Clinical Signs, Symptoms or Conditions	A071102	Failure to Interrogate	24/10/2022	24/10/2022
									A071204	No Pacing	24/10/2022	24/10/2022
							F1901	Additional Surgery	A071102	Failure to Interrogate	24/10/2022	24/10/2022
									A071204	No Pacing	24/10/2022	24/10/2022
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A071102	Failure to Interrogate	24/10/2022	24/10/2022
									A071204	No Pacing	24/10/2022	24/10/2022
							F1901	Additional Surgery	A071102	Failure to Interrogate	24/10/2022	24/10/2022
									A071204	No Pacing	24/10/2022	24/10/2022
					CT2727	Active implant devices/ancillaries	E2403	No Clinical Signs, Symptoms or Conditions	A071102	Failure to Interrogate	24/10/2022	24/10/2022
									A071204	No Pacing	24/10/2022	24/10/2022
							F1901	Additional Surgery	A071102	Failure to Interrogate	24/10/2022	24/10/2022

									A071204	No Pacing	24/10/2022	24/10/2022
					CT2750	Implantable cardiac electrical stimulators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A071102	Failure to Interrogate	24/10/2022	24/10/2022
									A071204	No Pacing	24/10/2022	24/10/2022
							F1901	Additional Surgery	A071102	Failure to Interrogate	24/10/2022	24/10/2022
									A071204	No Pacing	24/10/2022	24/10/2022
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A071102	Failure to Interrogate	24/10/2022	24/10/2022
									A071204	No Pacing	24/10/2022	24/10/2022
							F1901	Additional Surgery	A071102	Failure to Interrogate	24/10/2022	24/10/2022
									A071204	No Pacing	24/10/2022	24/10/2022
					CT2985	Active	E2403	No Clinical Signs, Symptoms or Conditions	A071102	Failure to Interrogate	24/10/2022	24/10/2022
									A071204	No Pacing	24/10/2022	24/10/2022
							F1901	Additional Surgery	A071102	Failure to Interrogate	24/10/2022	24/10/2022
									A071204	No Pacing	24/10/2022	24/10/2022
					CT3001	Implantable metal	E2403	No Clinical Signs, Symptoms or Conditions	A071102	Failure to Interrogate	24/10/2022	24/10/2022
									A071204	No Pacing	24/10/2022	24/10/2022
							F1901	Additional Surgery	A071102	Failure to Interrogate	24/10/2022	24/10/2022
									A071204	No Pacing	24/10/2022	24/10/2022
					CT3002	Implantable synthetic polymer	E2403	No Clinical Signs, Symptoms or Conditions	A071102	Failure to Interrogate	24/10/2022	24/10/2022
									A071204	No Pacing	24/10/2022	24/10/2022
							F1901	Additional Surgery	A071102	Failure to Interrogate	24/10/2022	24/10/2022
									A071204	No Pacing	24/10/2022	24/10/2022
					CT3006	Implantable inorganic material	E2403	No Clinical Signs, Symptoms or Conditions	A071102	Failure to Interrogate	24/10/2022	24/10/2022
									A071204	No Pacing	24/10/2022	24/10/2022
							F1901	Additional Surgery	A071102	Failure to Interrogate	24/10/2022	24/10/2022
									A071204	No Pacing	24/10/2022	24/10/2022
					CT314	Electrically powered	E2403	No Clinical Signs, Symptoms or Conditions	A071102	Failure to Interrogate	24/10/2022	24/10/2022
									A071204	No Pacing	24/10/2022	24/10/2022
							F1901	Additional Surgery	A071102	Failure to Interrogate	24/10/2022	24/10/2022
									A071204	No Pacing	24/10/2022	24/10/2022
					CT321	Long-term surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A071102	Failure to Interrogate	24/10/2022	24/10/2022
									A071204	No Pacing	24/10/2022	24/10/2022
							F1901	Additional Surgery	A071102	Failure to Interrogate	24/10/2022	24/10/2022
									A071204	No Pacing	24/10/2022	24/10/2022
					CT449	Generators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A071102	Failure to Interrogate	24/10/2022	24/10/2022
									A071204	No Pacing	24/10/2022	24/10/2022
							F1901	Additional Surgery	A071102	Failure to Interrogate	24/10/2022	24/10/2022
									A071204	No Pacing	24/10/2022	24/10/2022
					CT981	Single use	E2403	No Clinical Signs, Symptoms or Conditions	A071102	Failure to Interrogate	24/10/2022	24/10/2022
									A071204	No Pacing	24/10/2022	24/10/2022
							F1901	Additional Surgery	A071102	Failure to Interrogate	24/10/2022	24/10/2022
									A071204	No Pacing	24/10/2022	24/10/2022
					CT983	Surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A071102	Failure to Interrogate	24/10/2022	24/10/2022
									A071204	No Pacing	24/10/2022	24/10/2022
							F1901	Additional Surgery	A071102	Failure to Interrogate	24/10/2022	24/10/2022
									A071204	No Pacing	24/10/2022	24/10/2022
11570	2022	Unanticipated serious deterioration in state of health	60789	Intracardiac pacemaker	CT130	Catheters and associated devices	F	Annex F	A0101	Patient-Device Incompatibility		
					CT145	Cardiovascular devices	F	Annex F	A0101	Patient-Device Incompatibility		
					CT1684	Cardiac pacemakers and associated devices	F	Annex F	A0101	Patient-Device Incompatibility		
					CT1685	Cardiac pacemakers	F	Annex F	A0101	Patient-Device Incompatibility		
					CT233	Surgical	F	Annex F	A0101	Patient-Device Incompatibility		
					CT2406	Implantable	F	Annex F	A0101	Patient-Device Incompatibility		
					CT2407	Active implantable	F	Annex F	A0101	Patient-Device Incompatibility		
					CT257	Implantable cardiac pacemakers	F	Annex F	A0101	Patient-Device Incompatibility		
					CT2727	Active implant devices/ancillaries	F	Annex F	A0101	Patient-Device Incompatibility		
					CT2750	Implantable cardiac electrical stimulators and associated devices	F	Annex F	A0101	Patient-Device Incompatibility		
					CT2818	Catheterization/Cannulation/Fluid path devices	F	Annex F	A0101	Patient-Device Incompatibility		
					CT2866	Implantable cardiac electrical stimulators	F	Annex F	A0101	Patient-Device Incompatibility		
					CT2985	Active	F	Annex F	A0101	Patient-Device Incompatibility		
					CT3001	Implantable metal	F	Annex F	A0101	Patient-Device Incompatibility		
					CT3002	Implantable synthetic polymer	F	Annex F	A0101	Patient-Device Incompatibility		
					CT3006	Implantable inorganic material	F	Annex F	A0101	Patient-Device Incompatibility		
					CT314	Electrically powered	F	Annex F	A0101	Patient-Device Incompatibility		
					CT321	Long-term surgical invasive	F	Annex F	A0101	Patient-Device Incompatibility		
					CT981	Single use	F	Annex F	A0101	Patient-Device Incompatibility		
					CT983	Surgical invasive	F	Annex F	A0101	Patient-Device Incompatibility		
11571	2022	All other reportable incidents	47263	Cardiac resynchronization therapy implantable pacemaker	CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A0908	Incorrect, Inadequate or Imprecise Result or Readings		
							F26	No Health Consequences or Impact	A0908	Incorrect, Inadequate or Imprecise Result or Readings		
					CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A0908	Incorrect, Inadequate or Imprecise Result or Readings		
							F26	No Health Consequences or Impact	A0908	Incorrect, Inadequate or Imprecise Result or Readings		
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A0908	Incorrect, Inadequate or Imprecise Result or Readings		
							F26	No Health Consequences or Impact	A0908	Incorrect, Inadequate or Imprecise Result or Readings		
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A0908	Incorrect, Inadequate or Imprecise Result or Readings		
							F26	No Health Consequences or Impact	A0908	Incorrect, Inadequate or Imprecise Result or Readings		
11572	2022	All other reportable incidents	17882	Physiologic-monitoring defibrillation system	CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A07	Electrical /Electronic Property Problem		
							F27	Problem Identified During Non-Clinical Procedure	A07	Electrical /Electronic Property Problem		
11573	2022	All other reportable incidents	35852	Single-chamber implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A072201	High Impedance	04/11/2022	04/11/2022
									A0908	Incorrect, Inadequate or Imprecise Result or Readings	04/11/2022	04/11/2022
							F08	Hospitalization or Prolonged Hospitalization	A072201	High Impedance	04/11/2022	04/11/2022
									A0908	Incorrect, Inadequate or Imprecise Result or Readings	04/11/2022	04/11/2022

							F1905	Device Revision or Replacement	A072201	High Impedance	04/11/2022	04/11/2022
									A0908	Incorrect, Inadequate or Imprecise Result or Readings	04/11/2022	04/11/2022
11574	2022	All other reportable incidents	36241	Pacemaker/defibrillator lead	CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A072201	High Impedance		
							F08	Hospitalization or Prolonged Hospitalization	A072201	High Impedance		
							F1905	Device Revision or Replacement	A072201	High Impedance		
11575	2022	All other reportable incidents	35223	Endocardial/interventricular septal pacing lead	CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A0701	Capturing Problem		
									A070102	High Capture Threshold		
							F12	Serious Injury/ Illness/ Impairment	A0701	Capturing Problem		
									A070102	High Capture Threshold		
11576	2022	Unanticipated serious deterioration in state of health	35852	Single-chamber implantable defibrillator	CT2866	Implantable cardiac electrical stimulators			A27	Appropriate Device Problem Term/Code not Available		
11577	2022	Unanticipated serious deterioration in state of health	60789	Intracardiac pacemaker	CT130	Catheters and associated devices			A26	Insufficient Device Problem Information		
					CT145	Cardiovascular devices			A26	Insufficient Device Problem Information		
					CT1684	Cardiac pacemakers and associated devices			A26	Insufficient Device Problem Information		
					CT1685	Cardiac pacemakers			A26	Insufficient Device Problem Information		
					CT233	Surgical			A26	Insufficient Device Problem Information		
					CT2406	Implantable			A26	Insufficient Device Problem Information		
					CT2407	Active implantable			A26	Insufficient Device Problem Information		
					CT257	Implantable cardiac pacemakers			A26	Insufficient Device Problem Information		
					CT2727	Active implant devices/ancillaries			A26	Insufficient Device Problem Information		
					CT2750	Implantable cardiac electrical stimulators and associated devices			A26	Insufficient Device Problem Information		
					CT2818	Catheterization/Cannulation/Fluid path devices			A26	Insufficient Device Problem Information		
					CT2866	Implantable cardiac electrical stimulators			A26	Insufficient Device Problem Information		
					CT2985	Active			A26	Insufficient Device Problem Information		
					CT3001	Implantable metal			A26	Insufficient Device Problem Information		
					CT3002	Implantable synthetic polymer			A26	Insufficient Device Problem Information		
					CT3006	Implantable inorganic material			A26	Insufficient Device Problem Information		
					CT314	Electrically powered			A26	Insufficient Device Problem Information		
					CT321	Long-term surgical invasive			A26	Insufficient Device Problem Information		
					CT981	Single use			A26	Insufficient Device Problem Information		
					CT983	Surgical invasive			A26	Insufficient Device Problem Information		
11578	2022	Unanticipated serious deterioration in state of health	60789	Intracardiac pacemaker	CT130	Catheters and associated devices			A0709	Device Sensing Problem		
					CT145	Cardiovascular devices			A0709	Device Sensing Problem		
					CT1684	Cardiac pacemakers and associated devices			A0709	Device Sensing Problem		
					CT1685	Cardiac pacemakers			A0709	Device Sensing Problem		
					CT233	Surgical			A0709	Device Sensing Problem		
					CT2406	Implantable			A0709	Device Sensing Problem		
					CT2407	Active implantable			A0709	Device Sensing Problem		
					CT257	Implantable cardiac pacemakers			A0709	Device Sensing Problem		
					CT2727	Active implant devices/ancillaries			A0709	Device Sensing Problem		
					CT2750	Implantable cardiac electrical stimulators and associated devices			A0709	Device Sensing Problem		
					CT2818	Catheterization/Cannulation/Fluid path devices			A0709	Device Sensing Problem		
					CT2866	Implantable cardiac electrical stimulators			A0709	Device Sensing Problem		
					CT2985	Active			A0709	Device Sensing Problem		
					CT3001	Implantable metal			A0709	Device Sensing Problem		
					CT3002	Implantable synthetic polymer			A0709	Device Sensing Problem		
					CT3006	Implantable inorganic material			A0709	Device Sensing Problem		
					CT314	Electrically powered			A0709	Device Sensing Problem		
					CT321	Long-term surgical invasive			A0709	Device Sensing Problem		
					CT981	Single use			A0709	Device Sensing Problem		
					CT983	Surgical invasive			A0709	Device Sensing Problem		
11579	2022	All other reportable incidents	47265	Dual-chamber implantable pacemaker, rate-responsive	CT141	Electrodes/Leads and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A071102	Failure to Interrogate	10/11/2022	10/11/2022
							F1901	Additional Surgery	A071102	Failure to Interrogate	10/11/2022	10/11/2022
					CT145	Cardiovascular devices	E2403	No Clinical Signs, Symptoms or Conditions	A071102	Failure to Interrogate	10/11/2022	10/11/2022
							F1901	Additional Surgery	A071102	Failure to Interrogate	10/11/2022	10/11/2022
					CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A071102	Failure to Interrogate	10/11/2022	10/11/2022
							F1901	Additional Surgery	A071102	Failure to Interrogate	10/11/2022	10/11/2022
					CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A071102	Failure to Interrogate	10/11/2022	10/11/2022
							F1901	Additional Surgery	A071102	Failure to Interrogate	10/11/2022	10/11/2022
					CT233	Surgical	E2403	No Clinical Signs, Symptoms or Conditions	A071102	Failure to Interrogate	10/11/2022	10/11/2022
							F1901	Additional Surgery	A071102	Failure to Interrogate	10/11/2022	10/11/2022
					CT2406	Implantable	E2403	No Clinical Signs, Symptoms or Conditions	A071102	Failure to Interrogate	10/11/2022	10/11/2022
							F1901	Additional Surgery	A071102	Failure to Interrogate	10/11/2022	10/11/2022
					CT2407	Active implantable	E2403	No Clinical Signs, Symptoms or Conditions	A071102	Failure to Interrogate	10/11/2022	10/11/2022
							F1901	Additional Surgery	A071102	Failure to Interrogate	10/11/2022	10/11/2022
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A071102	Failure to Interrogate	10/11/2022	10/11/2022
							F1901	Additional Surgery	A071102	Failure to Interrogate	10/11/2022	10/11/2022
					CT2727	Active implant devices/ancillaries	E2403	No Clinical Signs, Symptoms or Conditions	A071102	Failure to Interrogate	10/11/2022	10/11/2022
							F1901	Additional Surgery	A071102	Failure to Interrogate	10/11/2022	10/11/2022
					CT2750	Implantable cardiac electrical stimulators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A071102	Failure to Interrogate	10/11/2022	10/11/2022
							F1901	Additional Surgery	A071102	Failure to Interrogate	10/11/2022	10/11/2022
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A071102	Failure to Interrogate	10/11/2022	10/11/2022
							F1901	Additional Surgery	A071102	Failure to Interrogate	10/11/2022	10/11/2022
					CT2985	Active	E2403	No Clinical Signs, Symptoms or Conditions	A071102	Failure to Interrogate	10/11/2022	10/11/2022
							F1901	Additional Surgery	A071102	Failure to Interrogate	10/11/2022	10/11/2022
					CT3001	Implantable metal	E2403	No Clinical Signs, Symptoms or Conditions	A071102	Failure to Interrogate	10/11/2022	10/11/2022
							F1901	Additional Surgery	A071102	Failure to Interrogate	10/11/2022	10/11/2022
					CT3002	Implantable synthetic polymer	E2403	No Clinical Signs, Symptoms or Conditions	A071102	Failure to Interrogate	10/11/2022	10/11/2022
							F1901	Additional Surgery	A071102	Failure to Interrogate	10/11/2022	10/11/2022
					CT3006	Implantable inorganic material	E2403	No Clinical Signs, Symptoms or Conditions	A071102	Failure to Interrogate	10/11/2022	10/11/2022
							F1901	Additional Surgery	A071102	Failure to Interrogate	10/11/2022	10/11/2022
					CT314	Electrically powered	E2403	No Clinical Signs, Symptoms or Conditions	A071102	Failure to Interrogate	10/11/2022	10/11/2022
							F1901	Additional Surgery	A071102	Failure to Interrogate	10/11/2022	10/11/2022
					CT321	Long-term surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A071102	Failure to Interrogate	10/11/2022	10/11/2022
							F1901	Additional Surgery	A071102	Failure to Interrogate	10/11/2022	10/11/2022
					CT449	Generators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A071102	Failure to Interrogate	10/11/2022	10/11/2022

							F1901	Additional Surgery	A071102	Failure to Interrogate	10/11/2022	10/11/2022
					CT981	Single use	E2403	No Clinical Signs, Symptoms or Conditions	A071102	Failure to Interrogate	10/11/2022	10/11/2022
							F1901	Additional Surgery	A071102	Failure to Interrogate	10/11/2022	10/11/2022
					CT983	Surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A071102	Failure to Interrogate	10/11/2022	10/11/2022
							F1901	Additional Surgery	A071102	Failure to Interrogate	10/11/2022	10/11/2022
11580	2022	All other reportable incidents	17882	Physiologic-monitoring defibrillation system	CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A1603	Failure of Device to Self-Test		
							F27	Problem Identified During Non-Clinical Procedure	A1603	Failure of Device to Self-Test		
11581	2022		35223	Endocardial/interventricular septal pacing lead	CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070909	Oversensing	04/11/2022	04/11/2022
							F1908	Prolonged Surgery	A070909	Oversensing	04/11/2022	04/11/2022
11582	2022	All other reportable incidents	17882	Physiologic-monitoring defibrillation system	CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A1603	Failure of Device to Self-Test		
							F26	No Health Consequences or Impact	A1603	Failure of Device to Self-Test		
11583	2022	All other reportable incidents	35223	Endocardial/interventricular septal pacing lead	CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070909	Oversensing		
							F12	Serious Injury/ Illness/ Impairment	A070909	Oversensing		
							F23	Unexpected Medical Intervention	A070909	Oversensing		
11584	2022	All other reportable incidents	47265	Dual-chamber implantable pacemaker, rate-responsive	CT141	Electrodes/Leads and associated devices	E011903	Syncope/Fainting	A070910	Undersensing		
									A0712	Pacing Problem		
									A072201	High Impedance		
							F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A070910	Undersensing		
									A0712	Pacing Problem		
									A072201	High Impedance		
							F22	Unexpected Diagnostic Intervention	A070910	Undersensing		
									A0712	Pacing Problem		
									A072201	High Impedance		
					CT145	Cardiovascular devices	E011903	Syncope/Fainting	A070910	Undersensing		
									A0712	Pacing Problem		
									A072201	High Impedance		
							F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A070910	Undersensing		
									A0712	Pacing Problem		
									A072201	High Impedance		
							F22	Unexpected Diagnostic Intervention	A070910	Undersensing		
									A0712	Pacing Problem		
									A072201	High Impedance		
					CT1684	Cardiac pacemakers and associated devices	E011903	Syncope/Fainting	A070910	Undersensing		
									A0712	Pacing Problem		
									A072201	High Impedance		
							F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A070910	Undersensing		
									A0712	Pacing Problem		
									A072201	High Impedance		
							F22	Unexpected Diagnostic Intervention	A070910	Undersensing		
									A0712	Pacing Problem		
									A072201	High Impedance		
					CT1685	Cardiac pacemakers	E011903	Syncope/Fainting	A070910	Undersensing		
									A0712	Pacing Problem		
									A072201	High Impedance		
							F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A070910	Undersensing		
									A0712	Pacing Problem		
									A072201	High Impedance		
							F22	Unexpected Diagnostic Intervention	A070910	Undersensing		
									A0712	Pacing Problem		
									A072201	High Impedance		
					CT233	Surgical	E011903	Syncope/Fainting	A070910	Undersensing		
									A0712	Pacing Problem		
									A072201	High Impedance		
							F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A070910	Undersensing		
									A0712	Pacing Problem		
									A072201	High Impedance		
							F22	Unexpected Diagnostic Intervention	A070910	Undersensing		
									A0712	Pacing Problem		
									A072201	High Impedance		
					CT2406	Implantable	E011903	Syncope/Fainting	A070910	Undersensing		
									A0712	Pacing Problem		
									A072201	High Impedance		
							F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A070910	Undersensing		
									A0712	Pacing Problem		
									A072201	High Impedance		
							F22	Unexpected Diagnostic Intervention	A070910	Undersensing		
									A0712	Pacing Problem		
									A072201	High Impedance		
					CT2407	Active implantable	E011903	Syncope/Fainting	A070910	Undersensing		
									A0712	Pacing Problem		
									A072201	High Impedance		
							F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A070910	Undersensing		
									A0712	Pacing Problem		
									A072201	High Impedance		
							F22	Unexpected Diagnostic Intervention	A070910	Undersensing		
									A0712	Pacing Problem		
									A072201	High Impedance		
					CT257	Implantable cardiac pacemakers	E011903	Syncope/Fainting	A070910	Undersensing		
									A0712	Pacing Problem		
									A072201	High Impedance		
							F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A070910	Undersensing		
									A0712	Pacing Problem		
									A072201	High Impedance		
							F22	Unexpected Diagnostic Intervention	A070910	Undersensing		
									A0712	Pacing Problem		
									A072201	High Impedance		
					CT2727	Active implant devices/ancillaries	E011903	Syncope/Fainting	A070910	Undersensing		
									A0712	Pacing Problem		

								A072201	High Impedance		
					F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A070910	Undersensing			
							A0712	Pacing Problem			
							A072201	High Impedance			
					F22	Unexpected Diagnostic Intervention	A070910	Undersensing			
							A0712	Pacing Problem			
							A072201	High Impedance			
				CT2750	Implantable cardiac electrical stimulators and associated devices	E011903	Syncope/Fainting	A070910	Undersensing		
								A0712	Pacing Problem		
								A072201	High Impedance		
						F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A070910	Undersensing		
								A0712	Pacing Problem		
								A072201	High Impedance		
						F22	Unexpected Diagnostic Intervention	A070910	Undersensing		
								A0712	Pacing Problem		
								A072201	High Impedance		
				CT2866	Implantable cardiac electrical stimulators	E011903	Syncope/Fainting	A070910	Undersensing		
								A0712	Pacing Problem		
								A072201	High Impedance		
						F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A070910	Undersensing		
								A0712	Pacing Problem		
								A072201	High Impedance		
						F22	Unexpected Diagnostic Intervention	A070910	Undersensing		
								A0712	Pacing Problem		
								A072201	High Impedance		
				CT2985	Active	E011903	Syncope/Fainting	A070910	Undersensing		
								A0712	Pacing Problem		
								A072201	High Impedance		
						F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A070910	Undersensing		
								A0712	Pacing Problem		
								A072201	High Impedance		
						F22	Unexpected Diagnostic Intervention	A070910	Undersensing		
								A0712	Pacing Problem		
								A072201	High Impedance		
				CT3001	Implantable metal	E011903	Syncope/Fainting	A070910	Undersensing		
								A0712	Pacing Problem		
								A072201	High Impedance		
						F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A070910	Undersensing		
								A0712	Pacing Problem		
								A072201	High Impedance		
						F22	Unexpected Diagnostic Intervention	A070910	Undersensing		
								A0712	Pacing Problem		
								A072201	High Impedance		
				CT3002	Implantable synthetic polymer	E011903	Syncope/Fainting	A070910	Undersensing		
								A0712	Pacing Problem		
								A072201	High Impedance		
						F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A070910	Undersensing		
								A0712	Pacing Problem		
								A072201	High Impedance		
						F22	Unexpected Diagnostic Intervention	A070910	Undersensing		
								A0712	Pacing Problem		
								A072201	High Impedance		
				CT3006	Implantable inorganic material	E011903	Syncope/Fainting	A070910	Undersensing		
								A0712	Pacing Problem		
								A072201	High Impedance		
						F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A070910	Undersensing		
								A0712	Pacing Problem		
								A072201	High Impedance		
						F22	Unexpected Diagnostic Intervention	A070910	Undersensing		
								A0712	Pacing Problem		
								A072201	High Impedance		
				CT314	Electrically powered	E011903	Syncope/Fainting	A070910	Undersensing		
								A0712	Pacing Problem		
								A072201	High Impedance		
						F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A070910	Undersensing		
								A0712	Pacing Problem		
								A072201	High Impedance		
						F22	Unexpected Diagnostic Intervention	A070910	Undersensing		
								A0712	Pacing Problem		
								A072201	High Impedance		
				CT321	Long-term surgical invasive	E011903	Syncope/Fainting	A070910	Undersensing		
								A0712	Pacing Problem		
								A072201	High Impedance		
						F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A070910	Undersensing		
								A0712	Pacing Problem		
								A072201	High Impedance		
						F22	Unexpected Diagnostic Intervention	A070910	Undersensing		
								A0712	Pacing Problem		
								A072201	High Impedance		
				CT449	Generators and associated devices	E011903	Syncope/Fainting	A070910	Undersensing		
								A0712	Pacing Problem		
								A072201	High Impedance		
						F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A070910	Undersensing		
								A0712	Pacing Problem		
								A072201	High Impedance		
						F22	Unexpected Diagnostic Intervention	A070910	Undersensing		
								A0712	Pacing Problem		
								A072201	High Impedance		

					CT981	Single use	E011903	Syncope/Fainting	A070910	Undersensing		
									A0712	Pacing Problem		
									A072201	High Impedance		
							F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A070910	Undersensing		
									A0712	Pacing Problem		
									A072201	High Impedance		
							F22	Unexpected Diagnostic Intervention	A070910	Undersensing		
									A0712	Pacing Problem		
									A072201	High Impedance		
					CT983	Surgical invasive	E011903	Syncope/Fainting	A070910	Undersensing		
									A0712	Pacing Problem		
									A072201	High Impedance		
							F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A070910	Undersensing		
									A0712	Pacing Problem		
									A072201	High Impedance		
							F22	Unexpected Diagnostic Intervention	A070910	Undersensing		
									A0712	Pacing Problem		
									A072201	High Impedance		
11585	2022	All other reportable incidents	47265	Dual-chamber implantable pacemaker, rate-responsive	CT141	Electrodes/Leads and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	02/11/2022	02/11/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	02/11/2022	02/11/2022
					CT145	Cardiovascular devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	02/11/2022	02/11/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	02/11/2022	02/11/2022
					CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	02/11/2022	02/11/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	02/11/2022	02/11/2022
					CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	02/11/2022	02/11/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	02/11/2022	02/11/2022
					CT233	Surgical	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	02/11/2022	02/11/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	02/11/2022	02/11/2022
					CT2406	Implantable	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	02/11/2022	02/11/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	02/11/2022	02/11/2022
					CT2407	Active implantable	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	02/11/2022	02/11/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	02/11/2022	02/11/2022
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	02/11/2022	02/11/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	02/11/2022	02/11/2022
					CT2727	Active implant devices/ancillaries	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	02/11/2022	02/11/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	02/11/2022	02/11/2022
					CT2750	Implantable cardiac electrical stimulators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	02/11/2022	02/11/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	02/11/2022	02/11/2022
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	02/11/2022	02/11/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	02/11/2022	02/11/2022
					CT2985	Active	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	02/11/2022	02/11/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	02/11/2022	02/11/2022
					CT3001	Implantable metal	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	02/11/2022	02/11/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	02/11/2022	02/11/2022
					CT3002	Implantable synthetic polymer	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	02/11/2022	02/11/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	02/11/2022	02/11/2022
					CT3006	Implantable inorganic material	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	02/11/2022	02/11/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	02/11/2022	02/11/2022
					CT314	Electrically powered	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	02/11/2022	02/11/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	02/11/2022	02/11/2022
					CT321	Long-term surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	02/11/2022	02/11/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	02/11/2022	02/11/2022
					CT449	Generators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	02/11/2022	02/11/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	02/11/2022	02/11/2022
					CT981	Single use	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	02/11/2022	02/11/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	02/11/2022	02/11/2022
					CT983	Surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	02/11/2022	02/11/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	02/11/2022	02/11/2022
11586	2022		47265	Dual-chamber implantable pacemaker, rate-responsive	CT141	Electrodes/Leads and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	04/11/2022	04/11/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	04/11/2022	04/11/2022
					CT145	Cardiovascular devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	04/11/2022	04/11/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	04/11/2022	04/11/2022
					CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	04/11/2022	04/11/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	04/11/2022	04/11/2022
					CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	04/11/2022	04/11/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	04/11/2022	04/11/2022
					CT233	Surgical	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	04/11/2022	04/11/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	04/11/2022	04/11/2022
					CT2406	Implantable	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	04/11/2022	04/11/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	04/11/2022	04/11/2022
					CT2407	Active implantable	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	04/11/2022	04/11/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	04/11/2022	04/11/2022
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	04/11/2022	04/11/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	04/11/2022	04/11/2022
					CT2727	Active implant devices/ancillaries	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	04/11/2022	04/11/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	04/11/2022	04/11/2022
					CT2750	Implantable cardiac electrical stimulators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	04/11/2022	04/11/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	04/11/2022	04/11/2022
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	04/11/2022	04/11/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	04/11/2022	04/11/2022
					CT2985	Active	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	04/11/2022	04/11/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	04/11/2022	04/11/2022
					CT3001	Implantable metal	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	04/11/2022	04/11/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	04/11/2022	04/11/2022
					CT3002	Implantable synthetic polymer	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	04/11/2022	04/11/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	04/11/2022	04/11/2022
					CT3006	Implantable inorganic material	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	04/11/2022	04/11/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	04/11/2022	04/11/2022

					CT314	Electrically powered	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	04/11/2022	04/11/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	04/11/2022	04/11/2022
					CT321	Long-term surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	04/11/2022	04/11/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	04/11/2022	04/11/2022
					CT449	Generators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	04/11/2022	04/11/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	04/11/2022	04/11/2022
					CT981	Single use	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	04/11/2022	04/11/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	04/11/2022	04/11/2022
					CT983	Surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	04/11/2022	04/11/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	04/11/2022	04/11/2022
11587	2022	All other reportable incidents	37265	Dual-chamber implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E060110	Ventricular Fibrillation	A0709	Device Sensing Problem		
									A0712	Pacing Problem		
									A071301	Failure to Deliver Shock/Stimulation		
							F02	Death	A0709	Device Sensing Problem		
									A0712	Pacing Problem		
									A071301	Failure to Deliver Shock/Stimulation		
							F08	Hospitalization or Prolonged Hospitalization	A0709	Device Sensing Problem		
									A0712	Pacing Problem		
									A071301	Failure to Deliver Shock/Stimulation		
							F1203	Life Threatening Illness or Injury	A0709	Device Sensing Problem		
									A0712	Pacing Problem		
									A071301	Failure to Deliver Shock/Stimulation		
							F2306	Resuscitation	A0709	Device Sensing Problem		
									A0712	Pacing Problem		
									A071301	Failure to Deliver Shock/Stimulation		
11588	2022	All other reportable incidents	17882	Physiologic-monitoring defibrillation system	CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070803	Failure to Power Up		
							F27	Problem Identified During Non-Clinical Procedure	A070803	Failure to Power Up		
11589	2022	All other reportable incidents	17882	Physiologic-monitoring defibrillation system	CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070803	Failure to Power Up		
							F27	Problem Identified During Non-Clinical Procedure	A070803	Failure to Power Up		
11590	2022	All other reportable incidents	47263	Cardiac resynchronization therapy implantable pacemaker	CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A0908	Incorrect, Inadequate or Imprecise Result or Readings	24/10/2022	24/10/2022
							F08	Hospitalization or Prolonged Hospitalization	A0908	Incorrect, Inadequate or Imprecise Result or Readings	24/10/2022	24/10/2022
							F1905	Device Revision or Replacement	A0908	Incorrect, Inadequate or Imprecise Result or Readings	24/10/2022	24/10/2022
					CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A0908	Incorrect, Inadequate or Imprecise Result or Readings	24/10/2022	24/10/2022
							F08	Hospitalization or Prolonged Hospitalization	A0908	Incorrect, Inadequate or Imprecise Result or Readings	24/10/2022	24/10/2022
							F1905	Device Revision or Replacement	A0908	Incorrect, Inadequate or Imprecise Result or Readings	24/10/2022	24/10/2022
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A0908	Incorrect, Inadequate or Imprecise Result or Readings	24/10/2022	24/10/2022
							F08	Hospitalization or Prolonged Hospitalization	A0908	Incorrect, Inadequate or Imprecise Result or Readings	24/10/2022	24/10/2022
							F1905	Device Revision or Replacement	A0908	Incorrect, Inadequate or Imprecise Result or Readings	24/10/2022	24/10/2022
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A0908	Incorrect, Inadequate or Imprecise Result or Readings	24/10/2022	24/10/2022
							F08	Hospitalization or Prolonged Hospitalization	A0908	Incorrect, Inadequate or Imprecise Result or Readings	24/10/2022	24/10/2022
							F1905	Device Revision or Replacement	A0908	Incorrect, Inadequate or Imprecise Result or Readings	24/10/2022	24/10/2022
11591	2022		47265	Dual-chamber implantable pacemaker, rate-responsive	CT141	Electrodes/Leads and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	06/11/2022	06/11/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	06/11/2022	06/11/2022
					CT145	Cardiovascular devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	06/11/2022	06/11/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	06/11/2022	06/11/2022
					CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	06/11/2022	06/11/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	06/11/2022	06/11/2022
					CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	06/11/2022	06/11/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	06/11/2022	06/11/2022
					CT233	Surgical	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	06/11/2022	06/11/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	06/11/2022	06/11/2022
					CT2406	Implantable	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	06/11/2022	06/11/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	06/11/2022	06/11/2022
					CT2407	Active implantable	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	06/11/2022	06/11/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	06/11/2022	06/11/2022
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	06/11/2022	06/11/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	06/11/2022	06/11/2022
					CT2727	Active implant devices/ancillaries	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	06/11/2022	06/11/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	06/11/2022	06/11/2022
					CT2750	Implantable cardiac electrical stimulators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	06/11/2022	06/11/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	06/11/2022	06/11/2022
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	06/11/2022	06/11/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	06/11/2022	06/11/2022
					CT2985	Active	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	06/11/2022	06/11/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	06/11/2022	06/11/2022
					CT3001	Implantable metal	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	06/11/2022	06/11/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	06/11/2022	06/11/2022
					CT3002	Implantable synthetic polymer	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	06/11/2022	06/11/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	06/11/2022	06/11/2022
					CT3006	Implantable inorganic material	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	06/11/2022	06/11/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	06/11/2022	06/11/2022
					CT314	Electrically powered	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	06/11/2022	06/11/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	06/11/2022	06/11/2022
					CT321	Long-term surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	06/11/2022	06/11/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	06/11/2022	06/11/2022
					CT449	Generators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	06/11/2022	06/11/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	06/11/2022	06/11/2022
					CT981	Single use	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	06/11/2022	06/11/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	06/11/2022	06/11/2022

					CT983	Surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	06/11/2022	06/11/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	06/11/2022	06/11/2022
11592	2022		47265	Dual-chamber implantable pacemaker, rate-responsive	CT141	Electrodes/Leads and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	06/11/2022	06/11/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	06/11/2022	06/11/2022
					CT145	Cardiovascular devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	06/11/2022	06/11/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	06/11/2022	06/11/2022
					CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	06/11/2022	06/11/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	06/11/2022	06/11/2022
					CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	06/11/2022	06/11/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	06/11/2022	06/11/2022
					CT233	Surgical	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	06/11/2022	06/11/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	06/11/2022	06/11/2022
					CT2406	Implantable	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	06/11/2022	06/11/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	06/11/2022	06/11/2022
					CT2407	Active implantable	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	06/11/2022	06/11/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	06/11/2022	06/11/2022
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	06/11/2022	06/11/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	06/11/2022	06/11/2022
					CT2727	Active implant devices/ancillaries	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	06/11/2022	06/11/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	06/11/2022	06/11/2022
					CT2750	Implantable cardiac electrical stimulators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	06/11/2022	06/11/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	06/11/2022	06/11/2022
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	06/11/2022	06/11/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	06/11/2022	06/11/2022
					CT2985	Active	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	06/11/2022	06/11/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	06/11/2022	06/11/2022
					CT3001	Implantable metal	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	06/11/2022	06/11/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	06/11/2022	06/11/2022
					CT3002	Implantable synthetic polymer	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	06/11/2022	06/11/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	06/11/2022	06/11/2022
					CT3006	Implantable inorganic material	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	06/11/2022	06/11/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	06/11/2022	06/11/2022
					CT314	Electrically powered	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	06/11/2022	06/11/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	06/11/2022	06/11/2022
					CT321	Long-term surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	06/11/2022	06/11/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	06/11/2022	06/11/2022
					CT449	Generators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	06/11/2022	06/11/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	06/11/2022	06/11/2022
					CT981	Single use	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	06/11/2022	06/11/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	06/11/2022	06/11/2022
					CT983	Surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	06/11/2022	06/11/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	06/11/2022	06/11/2022
11593	2022	All other reportable incidents	47263	Cardiac resynchronization therapy implantable pacemaker	CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A0908	Incorrect, Inadequate or Imprecise Result or Readings	17/10/2022	17/10/2022
							F08	Hospitalization or Prolonged Hospitalization	A0908	Incorrect, Inadequate or Imprecise Result or Readings	17/10/2022	17/10/2022
							F1905	Device Revision or Replacement	A0908	Incorrect, Inadequate or Imprecise Result or Readings	17/10/2022	17/10/2022
					CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A0908	Incorrect, Inadequate or Imprecise Result or Readings	17/10/2022	17/10/2022
							F08	Hospitalization or Prolonged Hospitalization	A0908	Incorrect, Inadequate or Imprecise Result or Readings	17/10/2022	17/10/2022
							F1905	Device Revision or Replacement	A0908	Incorrect, Inadequate or Imprecise Result or Readings	17/10/2022	17/10/2022
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A0908	Incorrect, Inadequate or Imprecise Result or Readings	17/10/2022	17/10/2022
							F08	Hospitalization or Prolonged Hospitalization	A0908	Incorrect, Inadequate or Imprecise Result or Readings	17/10/2022	17/10/2022
							F1905	Device Revision or Replacement	A0908	Incorrect, Inadequate or Imprecise Result or Readings	17/10/2022	17/10/2022
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A0908	Incorrect, Inadequate or Imprecise Result or Readings	17/10/2022	17/10/2022
							F08	Hospitalization or Prolonged Hospitalization	A0908	Incorrect, Inadequate or Imprecise Result or Readings	17/10/2022	17/10/2022
							F1905	Device Revision or Replacement	A0908	Incorrect, Inadequate or Imprecise Result or Readings	17/10/2022	17/10/2022
11594	2022	All other reportable incidents	36241	Pacemaker/defibrillator lead	CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070102	High Capture Threshold		
							F08	Hospitalization or Prolonged Hospitalization	A070102	High Capture Threshold		
							F1905	Device Revision or Replacement	A070102	High Capture Threshold		
11595	2022	All other reportable incidents	47265	Dual-chamber implantable pacemaker, rate-responsive	CT141	Electrodes/Leads and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
					CT145	Cardiovascular devices	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
					CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
					CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
					CT233	Surgical	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
					CT2406	Implantable	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		

							F26	No Health Consequences or Impact	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
					CT2407	Active implantable	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
					CT2727	Active implant devices/ancillaries	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
					CT2750	Implantable cardiac electrical stimulators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
					CT2985	Active	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
					CT3001	Implantable metal	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
					CT3002	Implantable synthetic polymer	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
					CT3006	Implantable inorganic material	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
					CT314	Electrically powered	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
					CT321	Long-term surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
					CT449	Generators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
					CT981	Single use	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
					CT983	Surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
11596	2022	All other reportable incidents	35852	Single-chamber implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A070908	Failure to Sense	28/09/2022	28/09/2022
							F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A070908	Failure to Sense	28/09/2022	28/09/2022
							F1908	Prolonged Surgery	A070908	Failure to Sense	28/09/2022	28/09/2022
11597	2022	All other reportable incidents	47265	Dual-chamber implantable pacemaker, rate-responsive	CT141	Electrodes/Leads and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
					CT145	Cardiovascular devices	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
					CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
					CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
					CT233	Surgical	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
					CT2406	Implantable	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
					CT2407	Active implantable	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A0705	Battery Problem		

									A070504	Premature Discharge of Battery		
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
					CT2727	Active implant devices/ancillaries	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
					CT2750	Implantable cardiac electrical stimulators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
					CT2985	Active	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
					CT3001	Implantable metal	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
					CT3002	Implantable synthetic polymer	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
					CT3006	Implantable inorganic material	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
					CT314	Electrically powered	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
					CT321	Long-term surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
					CT449	Generators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
					CT981	Single use	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
					CT983	Surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
11598	2022	All other reportable incidents	35223	Endocardial/interventricular septal pacing lead	CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A26	Insufficient Device Problem Information	28/10/2022	28/10/2022
							F1901	Additional Surgery	A26	Insufficient Device Problem Information	28/10/2022	28/10/2022
11599	2022	All other reportable incidents	35223	Endocardial/interventricular septal pacing lead	CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A0401	Break		
									A070909	Oversensing		
									A072202	Low Impedance		
							F22	Unexpected Diagnostic Intervention	A0401	Break		
									A070909	Oversensing		
									A072202	Low Impedance		
11600	2022	All other reportable incidents	47265	Dual-chamber implantable pacemaker, rate-responsive	CT141	Electrodes/Leads and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	06/11/2022	06/11/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	06/11/2022	06/11/2022
					CT145	Cardiovascular devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	06/11/2022	06/11/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	06/11/2022	06/11/2022
					CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	06/11/2022	06/11/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	06/11/2022	06/11/2022
					CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	06/11/2022	06/11/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	06/11/2022	06/11/2022
					CT233	Surgical	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	06/11/2022	06/11/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	06/11/2022	06/11/2022
					CT2406	Implantable	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	06/11/2022	06/11/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	06/11/2022	06/11/2022
					CT2407	Active implantable	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	06/11/2022	06/11/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	06/11/2022	06/11/2022
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	06/11/2022	06/11/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	06/11/2022	06/11/2022
					CT2727	Active implant devices/ancillaries	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	06/11/2022	06/11/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	06/11/2022	06/11/2022
					CT2750	Implantable cardiac electrical stimulators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	06/11/2022	06/11/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	06/11/2022	06/11/2022
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	06/11/2022	06/11/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	06/11/2022	06/11/2022
					CT2985	Active	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	06/11/2022	06/11/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	06/11/2022	06/11/2022
					CT3001	Implantable metal	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	06/11/2022	06/11/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	06/11/2022	06/11/2022
					CT3002	Implantable synthetic polymer	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	06/11/2022	06/11/2022

						F1901	Additional Surgery	A24	No Apparent Device Problem	06/11/2022	06/11/2022
					CT3006	Implantable inorganic material	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	06/11/2022	06/11/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	06/11/2022	06/11/2022
					CT314	Electrically powered	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	06/11/2022	06/11/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	06/11/2022	06/11/2022
					CT321	Long-term surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	06/11/2022	06/11/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	06/11/2022	06/11/2022
					CT449	Generators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	06/11/2022	06/11/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	06/11/2022	06/11/2022
					CT981	Single use	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	06/11/2022	06/11/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	06/11/2022	06/11/2022
					CT983	Surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	06/11/2022	06/11/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	06/11/2022	06/11/2022
11601	2022	All other reportable incidents	35223	Endocardial/interventricular septal pacing lead	CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070909	Oversensing	04/11/2022	04/11/2022
							F1908	Prolonged Surgery	A070909	Oversensing	04/11/2022	04/11/2022
11602	2022	All other reportable incidents	35039	Epicardial pacing lead	CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070901	Decreased Sensitivity		
									A0722	Impedance Problem		
							F08	Hospitalization or Prolonged Hospitalization	A070901	Decreased Sensitivity		
									A0722	Impedance Problem		
							F12	Serious Injury/ Illness/ Impairment	A070901	Decreased Sensitivity		
									A0722	Impedance Problem		
							F1904	Device Repositioning	A070901	Decreased Sensitivity		
									A0722	Impedance Problem		
11603	2022		47265	Dual-chamber implantable pacemaker, rate-responsive	CT141	Electrodes/Leads and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	07/11/2022	07/11/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	07/11/2022	07/11/2022
					CT145	Cardiovascular devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	07/11/2022	07/11/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	07/11/2022	07/11/2022
					CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	07/11/2022	07/11/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	07/11/2022	07/11/2022
					CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	07/11/2022	07/11/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	07/11/2022	07/11/2022
					CT233	Surgical	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	07/11/2022	07/11/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	07/11/2022	07/11/2022
					CT2406	Implantable	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	07/11/2022	07/11/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	07/11/2022	07/11/2022
					CT2407	Active implantable	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	07/11/2022	07/11/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	07/11/2022	07/11/2022
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	07/11/2022	07/11/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	07/11/2022	07/11/2022
					CT2727	Active implant devices/ancillaries	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	07/11/2022	07/11/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	07/11/2022	07/11/2022
					CT2750	Implantable cardiac electrical stimulators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	07/11/2022	07/11/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	07/11/2022	07/11/2022
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	07/11/2022	07/11/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	07/11/2022	07/11/2022
					CT2985	Active	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	07/11/2022	07/11/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	07/11/2022	07/11/2022
					CT3001	Implantable metal	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	07/11/2022	07/11/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	07/11/2022	07/11/2022
					CT3002	Implantable synthetic polymer	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	07/11/2022	07/11/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	07/11/2022	07/11/2022
					CT3006	Implantable inorganic material	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	07/11/2022	07/11/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	07/11/2022	07/11/2022
					CT314	Electrically powered	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	07/11/2022	07/11/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	07/11/2022	07/11/2022
					CT321	Long-term surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	07/11/2022	07/11/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	07/11/2022	07/11/2022
					CT449	Generators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	07/11/2022	07/11/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	07/11/2022	07/11/2022
					CT981	Single use	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	07/11/2022	07/11/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	07/11/2022	07/11/2022
					CT983	Surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	07/11/2022	07/11/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	07/11/2022	07/11/2022
11604	2022	All other reportable incidents	36241	Pacemaker/defibrillator lead	CT1684	Cardiac pacemakers and associated devices	E011903	Syncope/Fainting	A070910	Undersensing		
									A0712	Pacing Problem		
									A072201	High Impedance		
							F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A070910	Undersensing		
									A0712	Pacing Problem		
									A072201	High Impedance		
							F22	Unexpected Diagnostic Intervention	A070910	Undersensing		
									A0712	Pacing Problem		
									A072201	High Impedance		
11605	2022	All other reportable incidents	35852	Single-chamber implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem	29/12/2022	29/12/2022
									A070504	Premature Discharge of Battery	29/12/2022	29/12/2022
							F08	Hospitalization or Prolonged Hospitalization	A0705	Battery Problem	29/12/2022	29/12/2022
									A070504	Premature Discharge of Battery	29/12/2022	29/12/2022
							F1905	Device Revision or Replacement	A0705	Battery Problem	29/12/2022	29/12/2022
									A070504	Premature Discharge of Battery	29/12/2022	29/12/2022
11606	2022	All other reportable incidents	47267	Single-chamber implantable pacemaker, rate-responsive	CT141	Electrodes/Leads and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery		
							F12	Serious Injury/ Illness/ Impairment	A070504	Premature Discharge of Battery		
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery		
					CT145	Cardiovascular devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery		
							F12	Serious Injury/ Illness/ Impairment	A070504	Premature Discharge of Battery		
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery		
					CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery		

							F12	Serious Injury/ Illness/ Impairment	A070504	Premature Discharge of Battery		
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery		
					CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery		
							F12	Serious Injury/ Illness/ Impairment	A070504	Premature Discharge of Battery		
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery		
					CT233	Surgical	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery		
							F12	Serious Injury/ Illness/ Impairment	A070504	Premature Discharge of Battery		
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery		
					CT2406	Implantable	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery		
							F12	Serious Injury/ Illness/ Impairment	A070504	Premature Discharge of Battery		
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery		
					CT2407	Active implantable	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery		
							F12	Serious Injury/ Illness/ Impairment	A070504	Premature Discharge of Battery		
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery		
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery		
							F12	Serious Injury/ Illness/ Impairment	A070504	Premature Discharge of Battery		
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery		
					CT2727	Active implant devices/ancillaries	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery		
							F12	Serious Injury/ Illness/ Impairment	A070504	Premature Discharge of Battery		
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery		
					CT2750	Implantable cardiac electrical stimulators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery		
							F12	Serious Injury/ Illness/ Impairment	A070504	Premature Discharge of Battery		
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery		
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery		
							F12	Serious Injury/ Illness/ Impairment	A070504	Premature Discharge of Battery		
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery		
					CT2985	Active	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery		
							F12	Serious Injury/ Illness/ Impairment	A070504	Premature Discharge of Battery		
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery		
					CT3001	Implantable metal	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery		
							F12	Serious Injury/ Illness/ Impairment	A070504	Premature Discharge of Battery		
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery		
					CT3002	Implantable synthetic polymer	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery		
							F12	Serious Injury/ Illness/ Impairment	A070504	Premature Discharge of Battery		
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery		
					CT3006	Implantable inorganic material	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery		
							F12	Serious Injury/ Illness/ Impairment	A070504	Premature Discharge of Battery		
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery		
					CT314	Electrically powered	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery		
							F12	Serious Injury/ Illness/ Impairment	A070504	Premature Discharge of Battery		
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery		
					CT321	Long-term surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery		
							F12	Serious Injury/ Illness/ Impairment	A070504	Premature Discharge of Battery		
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery		
					CT449	Generators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery		
							F12	Serious Injury/ Illness/ Impairment	A070504	Premature Discharge of Battery		
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery		
					CT981	Single use	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery		
							F12	Serious Injury/ Illness/ Impairment	A070504	Premature Discharge of Battery		
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery		
					CT983	Surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery		
							F12	Serious Injury/ Illness/ Impairment	A070504	Premature Discharge of Battery		
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery		
11607	2022	All other reportable incidents	47265	Dual-chamber implantable pacemaker, rate-responsive	CT141	Electrodes/Leads and associated devices	E020204	Malaise	A070101	Failure to Capture		
									A0712	Pacing Problem		
							E2312	Fatigue	A070101	Failure to Capture		
									A0712	Pacing Problem		
							F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A070101	Failure to Capture		
									A0712	Pacing Problem		
							F15	Recognized Device or Procedural Complication	A070101	Failure to Capture		
									A0712	Pacing Problem		
					CT145	Cardiovascular devices	E020204	Malaise	A070101	Failure to Capture		
									A0712	Pacing Problem		
							E2312	Fatigue	A070101	Failure to Capture		
									A0712	Pacing Problem		
							F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A070101	Failure to Capture		
									A0712	Pacing Problem		
							F15	Recognized Device or Procedural Complication	A070101	Failure to Capture		
									A0712	Pacing Problem		
					CT1684	Cardiac pacemakers and associated devices	E020204	Malaise	A070101	Failure to Capture		
									A0712	Pacing Problem		

							E2312	Fatigue	A070101	Failure to Capture		
									A0712	Pacing Problem		
							F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A070101	Failure to Capture		
									A0712	Pacing Problem		
							F15	Recognized Device or Procedural Complication	A070101	Failure to Capture		
									A0712	Pacing Problem		
					CT1685	Cardiac pacemakers	E020204	Malaise	A070101	Failure to Capture		
									A0712	Pacing Problem		
							E2312	Fatigue	A070101	Failure to Capture		
									A0712	Pacing Problem		
							F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A070101	Failure to Capture		
									A0712	Pacing Problem		
							F15	Recognized Device or Procedural Complication	A070101	Failure to Capture		
									A0712	Pacing Problem		
					CT233	Surgical	E020204	Malaise	A070101	Failure to Capture		
									A0712	Pacing Problem		
							E2312	Fatigue	A070101	Failure to Capture		
									A0712	Pacing Problem		
							F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A070101	Failure to Capture		
									A0712	Pacing Problem		
							F15	Recognized Device or Procedural Complication	A070101	Failure to Capture		
									A0712	Pacing Problem		
					CT2406	Implantable	E020204	Malaise	A070101	Failure to Capture		
									A0712	Pacing Problem		
							E2312	Fatigue	A070101	Failure to Capture		
									A0712	Pacing Problem		
							F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A070101	Failure to Capture		
									A0712	Pacing Problem		
							F15	Recognized Device or Procedural Complication	A070101	Failure to Capture		
									A0712	Pacing Problem		
					CT2407	Active implantable	E020204	Malaise	A070101	Failure to Capture		
									A0712	Pacing Problem		
							E2312	Fatigue	A070101	Failure to Capture		
									A0712	Pacing Problem		
							F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A070101	Failure to Capture		
									A0712	Pacing Problem		
							F15	Recognized Device or Procedural Complication	A070101	Failure to Capture		
									A0712	Pacing Problem		
					CT257	Implantable cardiac pacemakers	E020204	Malaise	A070101	Failure to Capture		
									A0712	Pacing Problem		
							E2312	Fatigue	A070101	Failure to Capture		
									A0712	Pacing Problem		
							F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A070101	Failure to Capture		
									A0712	Pacing Problem		
							F15	Recognized Device or Procedural Complication	A070101	Failure to Capture		
									A0712	Pacing Problem		
					CT2727	Active implant devices/ancillaries	E020204	Malaise	A070101	Failure to Capture		
									A0712	Pacing Problem		
							E2312	Fatigue	A070101	Failure to Capture		
									A0712	Pacing Problem		
							F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A070101	Failure to Capture		
									A0712	Pacing Problem		
							F15	Recognized Device or Procedural Complication	A070101	Failure to Capture		
									A0712	Pacing Problem		
					CT2750	Implantable cardiac electrical stimulators and associated devices	E020204	Malaise	A070101	Failure to Capture		
									A0712	Pacing Problem		
							E2312	Fatigue	A070101	Failure to Capture		
									A0712	Pacing Problem		
							F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A070101	Failure to Capture		
									A0712	Pacing Problem		
							F15	Recognized Device or Procedural Complication	A070101	Failure to Capture		
									A0712	Pacing Problem		
					CT2866	Implantable cardiac electrical stimulators	E020204	Malaise	A070101	Failure to Capture		
									A0712	Pacing Problem		
							E2312	Fatigue	A070101	Failure to Capture		
									A0712	Pacing Problem		
							F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A070101	Failure to Capture		
									A0712	Pacing Problem		
							F15	Recognized Device or Procedural Complication	A070101	Failure to Capture		
									A0712	Pacing Problem		
					CT2985	Active	E020204	Malaise	A070101	Failure to Capture		
									A0712	Pacing Problem		
							E2312	Fatigue	A070101	Failure to Capture		
									A0712	Pacing Problem		
							F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A070101	Failure to Capture		
									A0712	Pacing Problem		
							F15	Recognized Device or Procedural Complication	A070101	Failure to Capture		
									A0712	Pacing Problem		
					CT3001	Implantable metal	E020204	Malaise	A070101	Failure to Capture		
									A0712	Pacing Problem		
							E2312	Fatigue	A070101	Failure to Capture		
									A0712	Pacing Problem		
							F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A070101	Failure to Capture		
									A0712	Pacing Problem		
							F15	Recognized Device or Procedural Complication	A070101	Failure to Capture		
									A0712	Pacing Problem		
					CT3002	Implantable synthetic polymer	E020204	Malaise	A070101	Failure to Capture		
									A0712	Pacing Problem		

							E2312	Fatigue	A070101	Failure to Capture		
									A0712	Pacing Problem		
							F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A070101	Failure to Capture		
									A0712	Pacing Problem		
							F15	Recognized Device or Procedural Complication	A070101	Failure to Capture		
									A0712	Pacing Problem		
					CT3006	Implantable inorganic material	E020204	Malaise	A070101	Failure to Capture		
									A0712	Pacing Problem		
							E2312	Fatigue	A070101	Failure to Capture		
									A0712	Pacing Problem		
							F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A070101	Failure to Capture		
									A0712	Pacing Problem		
							F15	Recognized Device or Procedural Complication	A070101	Failure to Capture		
									A0712	Pacing Problem		
					CT314	Electrically powered	E020204	Malaise	A070101	Failure to Capture		
									A0712	Pacing Problem		
							E2312	Fatigue	A070101	Failure to Capture		
									A0712	Pacing Problem		
							F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A070101	Failure to Capture		
									A0712	Pacing Problem		
							F15	Recognized Device or Procedural Complication	A070101	Failure to Capture		
									A0712	Pacing Problem		
					CT321	Long-term surgical invasive	E020204	Malaise	A070101	Failure to Capture		
									A0712	Pacing Problem		
							E2312	Fatigue	A070101	Failure to Capture		
									A0712	Pacing Problem		
							F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A070101	Failure to Capture		
									A0712	Pacing Problem		
							F15	Recognized Device or Procedural Complication	A070101	Failure to Capture		
									A0712	Pacing Problem		
					CT449	Generators and associated devices	E020204	Malaise	A070101	Failure to Capture		
									A0712	Pacing Problem		
							E2312	Fatigue	A070101	Failure to Capture		
									A0712	Pacing Problem		
							F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A070101	Failure to Capture		
									A0712	Pacing Problem		
							F15	Recognized Device or Procedural Complication	A070101	Failure to Capture		
									A0712	Pacing Problem		
					CT981	Single use	E020204	Malaise	A070101	Failure to Capture		
									A0712	Pacing Problem		
							E2312	Fatigue	A070101	Failure to Capture		
									A0712	Pacing Problem		
							F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A070101	Failure to Capture		
									A0712	Pacing Problem		
							F15	Recognized Device or Procedural Complication	A070101	Failure to Capture		
									A0712	Pacing Problem		
					CT983	Surgical invasive	E020204	Malaise	A070101	Failure to Capture		
									A0712	Pacing Problem		
							E2312	Fatigue	A070101	Failure to Capture		
									A0712	Pacing Problem		
							F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A070101	Failure to Capture		
									A0712	Pacing Problem		
							F15	Recognized Device or Procedural Complication	A070101	Failure to Capture		
									A0712	Pacing Problem		
11608	2022		47265	Dual-chamber implantable pacemaker, rate-responsive	CT141	Electrodes/Leads and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	06/11/2022	06/11/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	06/11/2022	06/11/2022
					CT145	Cardiovascular devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	06/11/2022	06/11/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	06/11/2022	06/11/2022
					CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	06/11/2022	06/11/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	06/11/2022	06/11/2022
					CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	06/11/2022	06/11/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	06/11/2022	06/11/2022
					CT233	Surgical	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	06/11/2022	06/11/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	06/11/2022	06/11/2022
					CT2406	Implantable	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	06/11/2022	06/11/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	06/11/2022	06/11/2022
					CT2407	Active implantable	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	06/11/2022	06/11/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	06/11/2022	06/11/2022
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	06/11/2022	06/11/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	06/11/2022	06/11/2022
					CT2727	Active implant devices/ancillaries	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	06/11/2022	06/11/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	06/11/2022	06/11/2022
					CT2750	Implantable cardiac electrical stimulators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	06/11/2022	06/11/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	06/11/2022	06/11/2022
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	06/11/2022	06/11/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	06/11/2022	06/11/2022
					CT2985	Active	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	06/11/2022	06/11/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	06/11/2022	06/11/2022
					CT3001	Implantable metal	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	06/11/2022	06/11/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	06/11/2022	06/11/2022
					CT3002	Implantable synthetic polymer	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	06/11/2022	06/11/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	06/11/2022	06/11/2022
					CT3006	Implantable inorganic material	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	06/11/2022	06/11/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	06/11/2022	06/11/2022
					CT314	Electrically powered	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	06/11/2022	06/11/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	06/11/2022	06/11/2022
					CT321	Long-term surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	06/11/2022	06/11/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	06/11/2022	06/11/2022

					CT449	Generators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	06/11/2022	06/11/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	06/11/2022	06/11/2022
					CT981	Single use	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	06/11/2022	06/11/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	06/11/2022	06/11/2022
					CT983	Surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	06/11/2022	06/11/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	06/11/2022	06/11/2022
11609	2022	All other reportable incidents	47265	Dual-chamber implantable pacemaker, rate-responsive	CT141	Electrodes/Leads and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	05/09/2022	05/09/2022
									A25	No Apparent Adverse Event	05/09/2022	05/09/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	05/09/2022	05/09/2022
									A25	No Apparent Adverse Event	05/09/2022	05/09/2022
					CT145	Cardiovascular devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	05/09/2022	05/09/2022
									A25	No Apparent Adverse Event	05/09/2022	05/09/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	05/09/2022	05/09/2022
									A25	No Apparent Adverse Event	05/09/2022	05/09/2022
					CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	05/09/2022	05/09/2022
									A25	No Apparent Adverse Event	05/09/2022	05/09/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	05/09/2022	05/09/2022
									A25	No Apparent Adverse Event	05/09/2022	05/09/2022
					CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	05/09/2022	05/09/2022
									A25	No Apparent Adverse Event	05/09/2022	05/09/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	05/09/2022	05/09/2022
									A25	No Apparent Adverse Event	05/09/2022	05/09/2022
					CT233	Surgical	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	05/09/2022	05/09/2022
									A25	No Apparent Adverse Event	05/09/2022	05/09/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	05/09/2022	05/09/2022
									A25	No Apparent Adverse Event	05/09/2022	05/09/2022
					CT2406	Implantable	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	05/09/2022	05/09/2022
									A25	No Apparent Adverse Event	05/09/2022	05/09/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	05/09/2022	05/09/2022
									A25	No Apparent Adverse Event	05/09/2022	05/09/2022
					CT2407	Active implantable	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	05/09/2022	05/09/2022
									A25	No Apparent Adverse Event	05/09/2022	05/09/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	05/09/2022	05/09/2022
									A25	No Apparent Adverse Event	05/09/2022	05/09/2022
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	05/09/2022	05/09/2022
									A25	No Apparent Adverse Event	05/09/2022	05/09/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	05/09/2022	05/09/2022
									A25	No Apparent Adverse Event	05/09/2022	05/09/2022
					CT2727	Active implant devices/ancillaries	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	05/09/2022	05/09/2022
									A25	No Apparent Adverse Event	05/09/2022	05/09/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	05/09/2022	05/09/2022
									A25	No Apparent Adverse Event	05/09/2022	05/09/2022
					CT2750	Implantable cardiac electrical stimulators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	05/09/2022	05/09/2022
									A25	No Apparent Adverse Event	05/09/2022	05/09/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	05/09/2022	05/09/2022
									A25	No Apparent Adverse Event	05/09/2022	05/09/2022
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	05/09/2022	05/09/2022
									A25	No Apparent Adverse Event	05/09/2022	05/09/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	05/09/2022	05/09/2022
									A25	No Apparent Adverse Event	05/09/2022	05/09/2022
					CT2985	Active	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	05/09/2022	05/09/2022
									A25	No Apparent Adverse Event	05/09/2022	05/09/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	05/09/2022	05/09/2022
									A25	No Apparent Adverse Event	05/09/2022	05/09/2022
					CT3001	Implantable metal	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	05/09/2022	05/09/2022
									A25	No Apparent Adverse Event	05/09/2022	05/09/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	05/09/2022	05/09/2022
									A25	No Apparent Adverse Event	05/09/2022	05/09/2022
					CT3002	Implantable synthetic polymer	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	05/09/2022	05/09/2022
									A25	No Apparent Adverse Event	05/09/2022	05/09/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	05/09/2022	05/09/2022
									A25	No Apparent Adverse Event	05/09/2022	05/09/2022
					CT3006	Implantable inorganic material	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	05/09/2022	05/09/2022
									A25	No Apparent Adverse Event	05/09/2022	05/09/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	05/09/2022	05/09/2022
									A25	No Apparent Adverse Event	05/09/2022	05/09/2022
					CT314	Electrically powered	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	05/09/2022	05/09/2022
									A25	No Apparent Adverse Event	05/09/2022	05/09/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	05/09/2022	05/09/2022
									A25	No Apparent Adverse Event	05/09/2022	05/09/2022
					CT321	Long-term surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	05/09/2022	05/09/2022
									A25	No Apparent Adverse Event	05/09/2022	05/09/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	05/09/2022	05/09/2022
									A25	No Apparent Adverse Event	05/09/2022	05/09/2022
					CT449	Generators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	05/09/2022	05/09/2022
									A25	No Apparent Adverse Event	05/09/2022	05/09/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	05/09/2022	05/09/2022
									A25	No Apparent Adverse Event	05/09/2022	05/09/2022
					CT981	Single use	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	05/09/2022	05/09/2022
									A25	No Apparent Adverse Event	05/09/2022	05/09/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	05/09/2022	05/09/2022
									A25	No Apparent Adverse Event	05/09/2022	05/09/2022
					CT983	Surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	05/09/2022	05/09/2022
									A25	No Apparent Adverse Event	05/09/2022	05/09/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	05/09/2022	05/09/2022
									A25	No Apparent Adverse Event	05/09/2022	05/09/2022
11610	2022	All other reportable incidents	47265	Dual-chamber implantable pacemaker, rate-responsive	CT141	Electrodes/Leads and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem	05/01/2023	05/01/2023
									A070504	Premature Discharge of Battery	05/01/2023	05/01/2023

		F08	Hospitalization or Prolonged Hospitalization	A0705	Battery Problem	05/01/2023	05/01/2023
				A070504	Premature Discharge of Battery	05/01/2023	05/01/2023
		F1905	Device Revision or Replacement	A0705	Battery Problem	05/01/2023	05/01/2023
				A070504	Premature Discharge of Battery	05/01/2023	05/01/2023
CT145	Cardiovascular devices	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem	05/01/2023	05/01/2023
				A070504	Premature Discharge of Battery	05/01/2023	05/01/2023
		F08	Hospitalization or Prolonged Hospitalization	A0705	Battery Problem	05/01/2023	05/01/2023
				A070504	Premature Discharge of Battery	05/01/2023	05/01/2023
		F1905	Device Revision or Replacement	A0705	Battery Problem	05/01/2023	05/01/2023
				A070504	Premature Discharge of Battery	05/01/2023	05/01/2023
CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem	05/01/2023	05/01/2023
				A070504	Premature Discharge of Battery	05/01/2023	05/01/2023
		F08	Hospitalization or Prolonged Hospitalization	A0705	Battery Problem	05/01/2023	05/01/2023
				A070504	Premature Discharge of Battery	05/01/2023	05/01/2023
		F1905	Device Revision or Replacement	A0705	Battery Problem	05/01/2023	05/01/2023
				A070504	Premature Discharge of Battery	05/01/2023	05/01/2023
CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem	05/01/2023	05/01/2023
				A070504	Premature Discharge of Battery	05/01/2023	05/01/2023
		F08	Hospitalization or Prolonged Hospitalization	A0705	Battery Problem	05/01/2023	05/01/2023
				A070504	Premature Discharge of Battery	05/01/2023	05/01/2023
		F1905	Device Revision or Replacement	A0705	Battery Problem	05/01/2023	05/01/2023
				A070504	Premature Discharge of Battery	05/01/2023	05/01/2023
CT233	Surgical	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem	05/01/2023	05/01/2023
				A070504	Premature Discharge of Battery	05/01/2023	05/01/2023
		F08	Hospitalization or Prolonged Hospitalization	A0705	Battery Problem	05/01/2023	05/01/2023
				A070504	Premature Discharge of Battery	05/01/2023	05/01/2023
		F1905	Device Revision or Replacement	A0705	Battery Problem	05/01/2023	05/01/2023
				A070504	Premature Discharge of Battery	05/01/2023	05/01/2023
CT2406	Implantable	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem	05/01/2023	05/01/2023
				A070504	Premature Discharge of Battery	05/01/2023	05/01/2023
		F08	Hospitalization or Prolonged Hospitalization	A0705	Battery Problem	05/01/2023	05/01/2023
				A070504	Premature Discharge of Battery	05/01/2023	05/01/2023
		F1905	Device Revision or Replacement	A0705	Battery Problem	05/01/2023	05/01/2023
				A070504	Premature Discharge of Battery	05/01/2023	05/01/2023
CT2407	Active implantable	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem	05/01/2023	05/01/2023
				A070504	Premature Discharge of Battery	05/01/2023	05/01/2023
		F08	Hospitalization or Prolonged Hospitalization	A0705	Battery Problem	05/01/2023	05/01/2023
				A070504	Premature Discharge of Battery	05/01/2023	05/01/2023
		F1905	Device Revision or Replacement	A0705	Battery Problem	05/01/2023	05/01/2023
				A070504	Premature Discharge of Battery	05/01/2023	05/01/2023
CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem	05/01/2023	05/01/2023
				A070504	Premature Discharge of Battery	05/01/2023	05/01/2023
		F08	Hospitalization or Prolonged Hospitalization	A0705	Battery Problem	05/01/2023	05/01/2023
				A070504	Premature Discharge of Battery	05/01/2023	05/01/2023
		F1905	Device Revision or Replacement	A0705	Battery Problem	05/01/2023	05/01/2023
				A070504	Premature Discharge of Battery	05/01/2023	05/01/2023
CT2727	Active implant devices/ancillaries	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem	05/01/2023	05/01/2023
				A070504	Premature Discharge of Battery	05/01/2023	05/01/2023
		F08	Hospitalization or Prolonged Hospitalization	A0705	Battery Problem	05/01/2023	05/01/2023
				A070504	Premature Discharge of Battery	05/01/2023	05/01/2023
		F1905	Device Revision or Replacement	A0705	Battery Problem	05/01/2023	05/01/2023
				A070504	Premature Discharge of Battery	05/01/2023	05/01/2023
CT2750	Implantable cardiac electrical stimulators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem	05/01/2023	05/01/2023
				A070504	Premature Discharge of Battery	05/01/2023	05/01/2023
		F08	Hospitalization or Prolonged Hospitalization	A0705	Battery Problem	05/01/2023	05/01/2023
				A070504	Premature Discharge of Battery	05/01/2023	05/01/2023
		F1905	Device Revision or Replacement	A0705	Battery Problem	05/01/2023	05/01/2023
				A070504	Premature Discharge of Battery	05/01/2023	05/01/2023
CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem	05/01/2023	05/01/2023
				A070504	Premature Discharge of Battery	05/01/2023	05/01/2023
		F08	Hospitalization or Prolonged Hospitalization	A0705	Battery Problem	05/01/2023	05/01/2023
				A070504	Premature Discharge of Battery	05/01/2023	05/01/2023
		F1905	Device Revision or Replacement	A0705	Battery Problem	05/01/2023	05/01/2023
				A070504	Premature Discharge of Battery	05/01/2023	05/01/2023
CT2985	Active	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem	05/01/2023	05/01/2023
				A070504	Premature Discharge of Battery	05/01/2023	05/01/2023
		F08	Hospitalization or Prolonged Hospitalization	A0705	Battery Problem	05/01/2023	05/01/2023
				A070504	Premature Discharge of Battery	05/01/2023	05/01/2023
		F1905	Device Revision or Replacement	A0705	Battery Problem	05/01/2023	05/01/2023
				A070504	Premature Discharge of Battery	05/01/2023	05/01/2023
CT3001	Implantable metal	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem	05/01/2023	05/01/2023
				A070504	Premature Discharge of Battery	05/01/2023	05/01/2023
		F08	Hospitalization or Prolonged Hospitalization	A0705	Battery Problem	05/01/2023	05/01/2023
				A070504	Premature Discharge of Battery	05/01/2023	05/01/2023
		F1905	Device Revision or Replacement	A0705	Battery Problem	05/01/2023	05/01/2023
				A070504	Premature Discharge of Battery	05/01/2023	05/01/2023
CT3002	Implantable synthetic polymer	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem	05/01/2023	05/01/2023
				A070504	Premature Discharge of Battery	05/01/2023	05/01/2023
		F08	Hospitalization or Prolonged Hospitalization	A0705	Battery Problem	05/01/2023	05/01/2023
				A070504	Premature Discharge of Battery	05/01/2023	05/01/2023
		F1905	Device Revision or Replacement	A0705	Battery Problem	05/01/2023	05/01/2023
				A070504	Premature Discharge of Battery	05/01/2023	05/01/2023
CT3006	Implantable inorganic material	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem	05/01/2023	05/01/2023
				A070504	Premature Discharge of Battery	05/01/2023	05/01/2023
		F08	Hospitalization or Prolonged Hospitalization	A0705	Battery Problem	05/01/2023	05/01/2023
				A070504	Premature Discharge of Battery	05/01/2023	05/01/2023
		F1905	Device Revision or Replacement	A0705	Battery Problem	05/01/2023	05/01/2023
				A070504	Premature Discharge of Battery	05/01/2023	05/01/2023

					CT314	Electrically powered	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem	05/01/2023	05/01/2023
									A070504	Premature Discharge of Battery	05/01/2023	05/01/2023
							F08	Hospitalization or Prolonged Hospitalization	A0705	Battery Problem	05/01/2023	05/01/2023
									A070504	Premature Discharge of Battery	05/01/2023	05/01/2023
							F1905	Device Revision or Replacement	A0705	Battery Problem	05/01/2023	05/01/2023
									A070504	Premature Discharge of Battery	05/01/2023	05/01/2023
					CT321	Long-term surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem	05/01/2023	05/01/2023
									A070504	Premature Discharge of Battery	05/01/2023	05/01/2023
							F08	Hospitalization or Prolonged Hospitalization	A0705	Battery Problem	05/01/2023	05/01/2023
									A070504	Premature Discharge of Battery	05/01/2023	05/01/2023
							F1905	Device Revision or Replacement	A0705	Battery Problem	05/01/2023	05/01/2023
									A070504	Premature Discharge of Battery	05/01/2023	05/01/2023
					CT449	Generators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem	05/01/2023	05/01/2023
									A070504	Premature Discharge of Battery	05/01/2023	05/01/2023
							F08	Hospitalization or Prolonged Hospitalization	A0705	Battery Problem	05/01/2023	05/01/2023
									A070504	Premature Discharge of Battery	05/01/2023	05/01/2023
							F1905	Device Revision or Replacement	A0705	Battery Problem	05/01/2023	05/01/2023
									A070504	Premature Discharge of Battery	05/01/2023	05/01/2023
					CT981	Single use	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem	05/01/2023	05/01/2023
									A070504	Premature Discharge of Battery	05/01/2023	05/01/2023
							F08	Hospitalization or Prolonged Hospitalization	A0705	Battery Problem	05/01/2023	05/01/2023
									A070504	Premature Discharge of Battery	05/01/2023	05/01/2023
							F1905	Device Revision or Replacement	A0705	Battery Problem	05/01/2023	05/01/2023
									A070504	Premature Discharge of Battery	05/01/2023	05/01/2023
					CT983	Surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem	05/01/2023	05/01/2023
									A070504	Premature Discharge of Battery	05/01/2023	05/01/2023
							F08	Hospitalization or Prolonged Hospitalization	A0705	Battery Problem	05/01/2023	05/01/2023
									A070504	Premature Discharge of Battery	05/01/2023	05/01/2023
							F1905	Device Revision or Replacement	A0705	Battery Problem	05/01/2023	05/01/2023
									A070504	Premature Discharge of Battery	05/01/2023	05/01/2023
11611	2022	All other reportable incidents			CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem		
							F26	No Health Consequences or Impact	A0705	Battery Problem		
11612	2022		47265	Dual-chamber implantable pacemaker, rate-responsive	CT141	Electrodes/Leads and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	07/11/2022	07/11/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	07/11/2022	07/11/2022
					CT145	Cardiovascular devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	07/11/2022	07/11/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	07/11/2022	07/11/2022
					CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	07/11/2022	07/11/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	07/11/2022	07/11/2022
					CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	07/11/2022	07/11/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	07/11/2022	07/11/2022
					CT233	Surgical	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	07/11/2022	07/11/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	07/11/2022	07/11/2022
					CT2406	Implantable	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	07/11/2022	07/11/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	07/11/2022	07/11/2022
					CT2407	Active implantable	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	07/11/2022	07/11/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	07/11/2022	07/11/2022
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	07/11/2022	07/11/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	07/11/2022	07/11/2022
					CT2727	Active implant devices/ancillaries	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	07/11/2022	07/11/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	07/11/2022	07/11/2022
					CT2750	Implantable cardiac electrical stimulators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	07/11/2022	07/11/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	07/11/2022	07/11/2022
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	07/11/2022	07/11/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	07/11/2022	07/11/2022
					CT2985	Active	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	07/11/2022	07/11/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	07/11/2022	07/11/2022
					CT3001	Implantable metal	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	07/11/2022	07/11/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	07/11/2022	07/11/2022
					CT3002	Implantable synthetic polymer	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	07/11/2022	07/11/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	07/11/2022	07/11/2022
					CT3006	Implantable inorganic material	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	07/11/2022	07/11/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	07/11/2022	07/11/2022
					CT314	Electrically powered	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	07/11/2022	07/11/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	07/11/2022	07/11/2022
					CT321	Long-term surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	07/11/2022	07/11/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	07/11/2022	07/11/2022
					CT449	Generators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	07/11/2022	07/11/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	07/11/2022	07/11/2022
					CT981	Single use	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	07/11/2022	07/11/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	07/11/2022	07/11/2022
					CT983	Surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	07/11/2022	07/11/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	07/11/2022	07/11/2022
11613	2022	All other reportable incidents	47265	Dual-chamber implantable pacemaker, rate-responsive	CT141	Electrodes/Leads and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A071102	Failure to Interrogate	10/11/2022	10/11/2022
									A071204	No Pacing	10/11/2022	10/11/2022
							F1901	Additional Surgery	A071102	Failure to Interrogate	10/11/2022	10/11/2022
									A071204	No Pacing	10/11/2022	10/11/2022
					CT145	Cardiovascular devices	E2403	No Clinical Signs, Symptoms or Conditions	A071102	Failure to Interrogate	10/11/2022	10/11/2022
									A071204	No Pacing	10/11/2022	10/11/2022
							F1901	Additional Surgery	A071102	Failure to Interrogate	10/11/2022	10/11/2022
									A071204	No Pacing	10/11/2022	10/11/2022
					CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A071102	Failure to Interrogate	10/11/2022	10/11/2022
									A071204	No Pacing	10/11/2022	10/11/2022
							F1901	Additional Surgery	A071102	Failure to Interrogate	10/11/2022	10/11/2022
									A071204	No Pacing	10/11/2022	10/11/2022
					CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A071102	Failure to Interrogate	10/11/2022	10/11/2022
									A071204	No Pacing	10/11/2022	10/11/2022
							F1901	Additional Surgery	A071102	Failure to Interrogate	10/11/2022	10/11/2022
									A071204	No Pacing	10/11/2022	10/11/2022

					CT233	Surgical	E2403	No Clinical Signs, Symptoms or Conditions	A071102	Failure to Interrogate	10/11/2022	10/11/2022
									A071204	No Pacing	10/11/2022	10/11/2022
							F1901	Additional Surgery	A071102	Failure to Interrogate	10/11/2022	10/11/2022
									A071204	No Pacing	10/11/2022	10/11/2022
					CT2406	Implantable	E2403	No Clinical Signs, Symptoms or Conditions	A071102	Failure to Interrogate	10/11/2022	10/11/2022
									A071204	No Pacing	10/11/2022	10/11/2022
							F1901	Additional Surgery	A071102	Failure to Interrogate	10/11/2022	10/11/2022
									A071204	No Pacing	10/11/2022	10/11/2022
					CT2407	Active implantable	E2403	No Clinical Signs, Symptoms or Conditions	A071102	Failure to Interrogate	10/11/2022	10/11/2022
									A071204	No Pacing	10/11/2022	10/11/2022
							F1901	Additional Surgery	A071102	Failure to Interrogate	10/11/2022	10/11/2022
									A071204	No Pacing	10/11/2022	10/11/2022
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A071102	Failure to Interrogate	10/11/2022	10/11/2022
									A071204	No Pacing	10/11/2022	10/11/2022
							F1901	Additional Surgery	A071102	Failure to Interrogate	10/11/2022	10/11/2022
									A071204	No Pacing	10/11/2022	10/11/2022
					CT2727	Active implant devices/ancillaries	E2403	No Clinical Signs, Symptoms or Conditions	A071102	Failure to Interrogate	10/11/2022	10/11/2022
									A071204	No Pacing	10/11/2022	10/11/2022
							F1901	Additional Surgery	A071102	Failure to Interrogate	10/11/2022	10/11/2022
									A071204	No Pacing	10/11/2022	10/11/2022
					CT2750	Implantable cardiac electrical stimulators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A071102	Failure to Interrogate	10/11/2022	10/11/2022
									A071204	No Pacing	10/11/2022	10/11/2022
							F1901	Additional Surgery	A071102	Failure to Interrogate	10/11/2022	10/11/2022
									A071204	No Pacing	10/11/2022	10/11/2022
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A071102	Failure to Interrogate	10/11/2022	10/11/2022
									A071204	No Pacing	10/11/2022	10/11/2022
							F1901	Additional Surgery	A071102	Failure to Interrogate	10/11/2022	10/11/2022
									A071204	No Pacing	10/11/2022	10/11/2022
					CT2985	Active	E2403	No Clinical Signs, Symptoms or Conditions	A071102	Failure to Interrogate	10/11/2022	10/11/2022
									A071204	No Pacing	10/11/2022	10/11/2022
							F1901	Additional Surgery	A071102	Failure to Interrogate	10/11/2022	10/11/2022
									A071204	No Pacing	10/11/2022	10/11/2022
					CT3001	Implantable metal	E2403	No Clinical Signs, Symptoms or Conditions	A071102	Failure to Interrogate	10/11/2022	10/11/2022
									A071204	No Pacing	10/11/2022	10/11/2022
							F1901	Additional Surgery	A071102	Failure to Interrogate	10/11/2022	10/11/2022
									A071204	No Pacing	10/11/2022	10/11/2022
					CT3002	Implantable synthetic polymer	E2403	No Clinical Signs, Symptoms or Conditions	A071102	Failure to Interrogate	10/11/2022	10/11/2022
									A071204	No Pacing	10/11/2022	10/11/2022
							F1901	Additional Surgery	A071102	Failure to Interrogate	10/11/2022	10/11/2022
									A071204	No Pacing	10/11/2022	10/11/2022
					CT3006	Implantable inorganic material	E2403	No Clinical Signs, Symptoms or Conditions	A071102	Failure to Interrogate	10/11/2022	10/11/2022
									A071204	No Pacing	10/11/2022	10/11/2022
							F1901	Additional Surgery	A071102	Failure to Interrogate	10/11/2022	10/11/2022
									A071204	No Pacing	10/11/2022	10/11/2022
					CT314	Electrically powered	E2403	No Clinical Signs, Symptoms or Conditions	A071102	Failure to Interrogate	10/11/2022	10/11/2022
									A071204	No Pacing	10/11/2022	10/11/2022
							F1901	Additional Surgery	A071102	Failure to Interrogate	10/11/2022	10/11/2022
									A071204	No Pacing	10/11/2022	10/11/2022
					CT321	Long-term surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A071102	Failure to Interrogate	10/11/2022	10/11/2022
									A071204	No Pacing	10/11/2022	10/11/2022
							F1901	Additional Surgery	A071102	Failure to Interrogate	10/11/2022	10/11/2022
									A071204	No Pacing	10/11/2022	10/11/2022
					CT449	Generators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A071102	Failure to Interrogate	10/11/2022	10/11/2022
									A071204	No Pacing	10/11/2022	10/11/2022
							F1901	Additional Surgery	A071102	Failure to Interrogate	10/11/2022	10/11/2022
									A071204	No Pacing	10/11/2022	10/11/2022
					CT981	Single use	E2403	No Clinical Signs, Symptoms or Conditions	A071102	Failure to Interrogate	10/11/2022	10/11/2022
									A071204	No Pacing	10/11/2022	10/11/2022
							F1901	Additional Surgery	A071102	Failure to Interrogate	10/11/2022	10/11/2022
									A071204	No Pacing	10/11/2022	10/11/2022
					CT983	Surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A071102	Failure to Interrogate	10/11/2022	10/11/2022
									A071204	No Pacing	10/11/2022	10/11/2022
							F1901	Additional Surgery	A071102	Failure to Interrogate	10/11/2022	10/11/2022
									A071204	No Pacing	10/11/2022	10/11/2022
11614	2022		47265	Dual-chamber implantable pacemaker, rate-responsive	CT141	Electrodes/Leads and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	17/11/2022	17/11/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	17/11/2022	17/11/2022
					CT145	Cardiovascular devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	17/11/2022	17/11/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	17/11/2022	17/11/2022
					CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	17/11/2022	17/11/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	17/11/2022	17/11/2022
					CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	17/11/2022	17/11/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	17/11/2022	17/11/2022
					CT233	Surgical	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	17/11/2022	17/11/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	17/11/2022	17/11/2022
					CT2406	Implantable	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	17/11/2022	17/11/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	17/11/2022	17/11/2022
					CT2407	Active implantable	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	17/11/2022	17/11/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	17/11/2022	17/11/2022
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	17/11/2022	17/11/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	17/11/2022	17/11/2022
					CT2727	Active implant devices/ancillaries	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	17/11/2022	17/11/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	17/11/2022	17/11/2022
					CT2750	Implantable cardiac electrical stimulators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	17/11/2022	17/11/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	17/11/2022	17/11/2022
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	17/11/2022	17/11/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	17/11/2022	17/11/2022
					CT2985	Active	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	17/11/2022	17/11/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	17/11/2022	17/11/2022

					CT3001	Implantable metal	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	17/11/2022	17/11/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	17/11/2022	17/11/2022
					CT3002	Implantable synthetic polymer	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	17/11/2022	17/11/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	17/11/2022	17/11/2022
					CT3006	Implantable inorganic material	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	17/11/2022	17/11/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	17/11/2022	17/11/2022
					CT314	Electrically powered	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	17/11/2022	17/11/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	17/11/2022	17/11/2022
					CT321	Long-term surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	17/11/2022	17/11/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	17/11/2022	17/11/2022
					CT449	Generators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	17/11/2022	17/11/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	17/11/2022	17/11/2022
					CT981	Single use	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	17/11/2022	17/11/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	17/11/2022	17/11/2022
					CT983	Surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	17/11/2022	17/11/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	17/11/2022	17/11/2022
11615	2022	All other reportable incidents	36241	Pacemaker/defibrillator lead	CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	15/12/2022	15/12/2022
									A0908	Incorrect, Inadequate or Imprecise Result or Readings	15/12/2022	15/12/2022
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	15/12/2022	15/12/2022
									A0908	Incorrect, Inadequate or Imprecise Result or Readings	15/12/2022	15/12/2022
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	15/12/2022	15/12/2022
									A0908	Incorrect, Inadequate or Imprecise Result or Readings	15/12/2022	15/12/2022
11616	2022	All other reportable incidents	47265	Dual-chamber implantable pacemaker, rate-responsive	CT141	Electrodes/Leads and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	12/04/2023	12/04/2023
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	12/04/2023	12/04/2023
							F1903	Device Explantation	A070504	Premature Discharge of Battery	12/04/2023	12/04/2023
					CT145	Cardiovascular devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	12/04/2023	12/04/2023
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	12/04/2023	12/04/2023
							F1903	Device Explantation	A070504	Premature Discharge of Battery	12/04/2023	12/04/2023
					CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	12/04/2023	12/04/2023
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	12/04/2023	12/04/2023
							F1903	Device Explantation	A070504	Premature Discharge of Battery	12/04/2023	12/04/2023
					CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	12/04/2023	12/04/2023
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	12/04/2023	12/04/2023
							F1903	Device Explantation	A070504	Premature Discharge of Battery	12/04/2023	12/04/2023
					CT233	Surgical	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	12/04/2023	12/04/2023
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	12/04/2023	12/04/2023
							F1903	Device Explantation	A070504	Premature Discharge of Battery	12/04/2023	12/04/2023
					CT2406	Implantable	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	12/04/2023	12/04/2023
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	12/04/2023	12/04/2023
							F1903	Device Explantation	A070504	Premature Discharge of Battery	12/04/2023	12/04/2023
					CT2407	Active implantable	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	12/04/2023	12/04/2023
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	12/04/2023	12/04/2023
							F1903	Device Explantation	A070504	Premature Discharge of Battery	12/04/2023	12/04/2023
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	12/04/2023	12/04/2023
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	12/04/2023	12/04/2023
							F1903	Device Explantation	A070504	Premature Discharge of Battery	12/04/2023	12/04/2023
					CT2727	Active implant devices/ancillaries	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	12/04/2023	12/04/2023
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	12/04/2023	12/04/2023
							F1903	Device Explantation	A070504	Premature Discharge of Battery	12/04/2023	12/04/2023
					CT2750	Implantable cardiac electrical stimulators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	12/04/2023	12/04/2023
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	12/04/2023	12/04/2023
							F1903	Device Explantation	A070504	Premature Discharge of Battery	12/04/2023	12/04/2023
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	12/04/2023	12/04/2023
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	12/04/2023	12/04/2023
							F1903	Device Explantation	A070504	Premature Discharge of Battery	12/04/2023	12/04/2023
					CT2985	Active	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	12/04/2023	12/04/2023
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	12/04/2023	12/04/2023
							F1903	Device Explantation	A070504	Premature Discharge of Battery	12/04/2023	12/04/2023
					CT3001	Implantable metal	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	12/04/2023	12/04/2023
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	12/04/2023	12/04/2023
							F1903	Device Explantation	A070504	Premature Discharge of Battery	12/04/2023	12/04/2023
					CT3002	Implantable synthetic polymer	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	12/04/2023	12/04/2023
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	12/04/2023	12/04/2023
							F1903	Device Explantation	A070504	Premature Discharge of Battery	12/04/2023	12/04/2023
					CT3006	Implantable inorganic material	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	12/04/2023	12/04/2023
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	12/04/2023	12/04/2023
							F1903	Device Explantation	A070504	Premature Discharge of Battery	12/04/2023	12/04/2023
					CT314	Electrically powered	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	12/04/2023	12/04/2023
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	12/04/2023	12/04/2023
							F1903	Device Explantation	A070504	Premature Discharge of Battery	12/04/2023	12/04/2023
					CT321	Long-term surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	12/04/2023	12/04/2023
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	12/04/2023	12/04/2023
							F1903	Device Explantation	A070504	Premature Discharge of Battery	12/04/2023	12/04/2023
					CT449	Generators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	12/04/2023	12/04/2023
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	12/04/2023	12/04/2023
							F1903	Device Explantation	A070504	Premature Discharge of Battery	12/04/2023	12/04/2023
					CT981	Single use	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	12/04/2023	12/04/2023
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	12/04/2023	12/04/2023
							F1903	Device Explantation	A070504	Premature Discharge of Battery	12/04/2023	12/04/2023
					CT983	Surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	12/04/2023	12/04/2023
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	12/04/2023	12/04/2023
							F1903	Device Explantation	A070504	Premature Discharge of Battery	12/04/2023	12/04/2023
11617	2022	All other reportable incidents	35223	Endocardial/interventricular septal pacing lead	CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070909	Oversensing		
									A090801	Signal Artifact/Noise		
							F26	No Health Consequences or Impact	A070909	Oversensing		

									A090801	Signal Artifact/Noise		
11618	2022	All other reportable incidents	36241	Pacemaker/defibrillator lead	CT1684	Cardiac pacemakers and associated devices	E2104	Electric Shock	A070909	Oversensing	24/11/2022	24/11/2022
									A071302	Inappropriate/Inadequate Shock/Stimulation	24/11/2022	24/11/2022
									A090801	Signal Artifact/Noise	24/11/2022	24/11/2022
									A1301	Failure to Read Input Signal	24/11/2022	24/11/2022
							F08	Hospitalization or Prolonged Hospitalization	A070909	Oversensing	24/11/2022	24/11/2022
									A071302	Inappropriate/Inadequate Shock/Stimulation	24/11/2022	24/11/2022
									A090801	Signal Artifact/Noise	24/11/2022	24/11/2022
									A1301	Failure to Read Input Signal	24/11/2022	24/11/2022
							F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A070909	Oversensing	24/11/2022	24/11/2022
									A071302	Inappropriate/Inadequate Shock/Stimulation	24/11/2022	24/11/2022
									A090801	Signal Artifact/Noise	24/11/2022	24/11/2022
									A1301	Failure to Read Input Signal	24/11/2022	24/11/2022
							F1905	Device Revision or Replacement	A070909	Oversensing	24/11/2022	24/11/2022
									A071302	Inappropriate/Inadequate Shock/Stimulation	24/11/2022	24/11/2022
									A090801	Signal Artifact/Noise	24/11/2022	24/11/2022
									A1301	Failure to Read Input Signal	24/11/2022	24/11/2022
11619	2022	All other reportable incidents	37265	Dual-chamber implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
11620	2022		47265	Dual-chamber implantable pacemaker, rate-responsive	CT141	Electrodes/Leads and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	10/11/2022	10/11/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	10/11/2022	10/11/2022
					CT145	Cardiovascular devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	10/11/2022	10/11/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	10/11/2022	10/11/2022
					CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	10/11/2022	10/11/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	10/11/2022	10/11/2022
					CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	10/11/2022	10/11/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	10/11/2022	10/11/2022
					CT233	Surgical	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	10/11/2022	10/11/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	10/11/2022	10/11/2022
					CT2406	Implantable	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	10/11/2022	10/11/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	10/11/2022	10/11/2022
					CT2407	Active implantable	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	10/11/2022	10/11/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	10/11/2022	10/11/2022
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	10/11/2022	10/11/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	10/11/2022	10/11/2022
					CT2727	Active implant devices/ancillaries	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	10/11/2022	10/11/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	10/11/2022	10/11/2022
					CT2750	Implantable cardiac electrical stimulators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	10/11/2022	10/11/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	10/11/2022	10/11/2022
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	10/11/2022	10/11/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	10/11/2022	10/11/2022
					CT2985	Active	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	10/11/2022	10/11/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	10/11/2022	10/11/2022
					CT3001	Implantable metal	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	10/11/2022	10/11/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	10/11/2022	10/11/2022
					CT3002	Implantable synthetic polymer	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	10/11/2022	10/11/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	10/11/2022	10/11/2022
					CT3006	Implantable inorganic material	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	10/11/2022	10/11/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	10/11/2022	10/11/2022
					CT314	Electrically powered	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	10/11/2022	10/11/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	10/11/2022	10/11/2022
					CT321	Long-term surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	10/11/2022	10/11/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	10/11/2022	10/11/2022
					CT449	Generators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	10/11/2022	10/11/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	10/11/2022	10/11/2022
					CT981	Single use	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	10/11/2022	10/11/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	10/11/2022	10/11/2022
					CT983	Surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	10/11/2022	10/11/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	10/11/2022	10/11/2022
11621	2022	All other reportable incidents	35852	Single-chamber implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
11622	2022	All other reportable incidents	47265	Dual-chamber implantable pacemaker, rate-responsive	CT141	Electrodes/Leads and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	13/10/2022	13/10/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	13/10/2022	13/10/2022
					CT145	Cardiovascular devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	13/10/2022	13/10/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	13/10/2022	13/10/2022
					CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	13/10/2022	13/10/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	13/10/2022	13/10/2022
					CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	13/10/2022	13/10/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	13/10/2022	13/10/2022
					CT233	Surgical	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	13/10/2022	13/10/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	13/10/2022	13/10/2022
					CT2406	Implantable	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	13/10/2022	13/10/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	13/10/2022	13/10/2022
					CT2407	Active implantable	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	13/10/2022	13/10/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	13/10/2022	13/10/2022
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	13/10/2022	13/10/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	13/10/2022	13/10/2022
					CT2727	Active implant devices/ancillaries	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	13/10/2022	13/10/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	13/10/2022	13/10/2022
					CT2750	Implantable cardiac electrical stimulators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	13/10/2022	13/10/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	13/10/2022	13/10/2022
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	13/10/2022	13/10/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	13/10/2022	13/10/2022
					CT2985	Active	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	13/10/2022	13/10/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	13/10/2022	13/10/2022
					CT3001	Implantable metal	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	13/10/2022	13/10/2022

							F1901	Additional Surgery	A24	No Apparent Device Problem	13/10/2022	13/10/2022
					CT3002	Implantable synthetic polymer	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	13/10/2022	13/10/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	13/10/2022	13/10/2022
					CT3006	Implantable inorganic material	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	13/10/2022	13/10/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	13/10/2022	13/10/2022
					CT314	Electrically powered	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	13/10/2022	13/10/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	13/10/2022	13/10/2022
					CT321	Long-term surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	13/10/2022	13/10/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	13/10/2022	13/10/2022
					CT449	Generators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	13/10/2022	13/10/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	13/10/2022	13/10/2022
					CT981	Single use	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	13/10/2022	13/10/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	13/10/2022	13/10/2022
					CT983	Surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	13/10/2022	13/10/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	13/10/2022	13/10/2022
11623	2022		47265	Dual-chamber implantable pacemaker, rate-responsive	CT141	Electrodes/Leads and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	16/11/2022	16/11/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	16/11/2022	16/11/2022
					CT145	Cardiovascular devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	16/11/2022	16/11/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	16/11/2022	16/11/2022
					CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	16/11/2022	16/11/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	16/11/2022	16/11/2022
					CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	16/11/2022	16/11/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	16/11/2022	16/11/2022
					CT233	Surgical	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	16/11/2022	16/11/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	16/11/2022	16/11/2022
					CT2406	Implantable	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	16/11/2022	16/11/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	16/11/2022	16/11/2022
					CT2407	Active implantable	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	16/11/2022	16/11/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	16/11/2022	16/11/2022
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	16/11/2022	16/11/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	16/11/2022	16/11/2022
					CT2727	Active implant devices/ancillaries	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	16/11/2022	16/11/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	16/11/2022	16/11/2022
					CT2750	Implantable cardiac electrical stimulators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	16/11/2022	16/11/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	16/11/2022	16/11/2022
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	16/11/2022	16/11/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	16/11/2022	16/11/2022
					CT2985	Active	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	16/11/2022	16/11/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	16/11/2022	16/11/2022
					CT3001	Implantable metal	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	16/11/2022	16/11/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	16/11/2022	16/11/2022
					CT3002	Implantable synthetic polymer	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	16/11/2022	16/11/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	16/11/2022	16/11/2022
					CT3006	Implantable inorganic material	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	16/11/2022	16/11/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	16/11/2022	16/11/2022
					CT314	Electrically powered	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	16/11/2022	16/11/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	16/11/2022	16/11/2022
					CT321	Long-term surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	16/11/2022	16/11/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	16/11/2022	16/11/2022
					CT449	Generators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	16/11/2022	16/11/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	16/11/2022	16/11/2022
					CT981	Single use	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	16/11/2022	16/11/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	16/11/2022	16/11/2022
					CT983	Surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	16/11/2022	16/11/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	16/11/2022	16/11/2022
11624	2022		47265	Dual-chamber implantable pacemaker, rate-responsive	CT141	Electrodes/Leads and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	14/11/2022	14/11/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	14/11/2022	14/11/2022
					CT145	Cardiovascular devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	14/11/2022	14/11/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	14/11/2022	14/11/2022
					CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	14/11/2022	14/11/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	14/11/2022	14/11/2022
					CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	14/11/2022	14/11/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	14/11/2022	14/11/2022
					CT233	Surgical	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	14/11/2022	14/11/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	14/11/2022	14/11/2022
					CT2406	Implantable	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	14/11/2022	14/11/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	14/11/2022	14/11/2022
					CT2407	Active implantable	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	14/11/2022	14/11/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	14/11/2022	14/11/2022
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	14/11/2022	14/11/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	14/11/2022	14/11/2022
					CT2727	Active implant devices/ancillaries	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	14/11/2022	14/11/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	14/11/2022	14/11/2022
					CT2750	Implantable cardiac electrical stimulators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	14/11/2022	14/11/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	14/11/2022	14/11/2022
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	14/11/2022	14/11/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	14/11/2022	14/11/2022
					CT2985	Active	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	14/11/2022	14/11/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	14/11/2022	14/11/2022
					CT3001	Implantable metal	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	14/11/2022	14/11/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	14/11/2022	14/11/2022
					CT3002	Implantable synthetic polymer	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	14/11/2022	14/11/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	14/11/2022	14/11/2022
					CT3006	Implantable inorganic material	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	14/11/2022	14/11/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	14/11/2022	14/11/2022
					CT314	Electrically powered	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	14/11/2022	14/11/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	14/11/2022	14/11/2022
					CT321	Long-term surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	14/11/2022	14/11/2022

						F1901	Additional Surgery	A24	No Apparent Device Problem	14/11/2022	14/11/2022
				CT449	Generators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	14/11/2022	14/11/2022
						F1901	Additional Surgery	A24	No Apparent Device Problem	14/11/2022	14/11/2022
				CT981	Single use	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	14/11/2022	14/11/2022
						F1901	Additional Surgery	A24	No Apparent Device Problem	14/11/2022	14/11/2022
				CT983	Surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	14/11/2022	14/11/2022
						F1901	Additional Surgery	A24	No Apparent Device Problem	14/11/2022	14/11/2022
11625	2022	All other reportable incidents	47265	Dual-chamber implantable pacemaker, rate-responsive	CT141	Electrodes/Leads and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	15/11/2022	15/11/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	15/11/2022	15/11/2022
					CT145	Cardiovascular devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	15/11/2022	15/11/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	15/11/2022	15/11/2022
					CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	15/11/2022	15/11/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	15/11/2022	15/11/2022
					CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	15/11/2022	15/11/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	15/11/2022	15/11/2022
					CT233	Surgical	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	15/11/2022	15/11/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	15/11/2022	15/11/2022
					CT2406	Implantable	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	15/11/2022	15/11/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	15/11/2022	15/11/2022
					CT2407	Active implantable	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	15/11/2022	15/11/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	15/11/2022	15/11/2022
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	15/11/2022	15/11/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	15/11/2022	15/11/2022
					CT2727	Active implant devices/ancillaries	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	15/11/2022	15/11/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	15/11/2022	15/11/2022
					CT2750	Implantable cardiac electrical stimulators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	15/11/2022	15/11/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	15/11/2022	15/11/2022
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	15/11/2022	15/11/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	15/11/2022	15/11/2022
					CT2985	Active	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	15/11/2022	15/11/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	15/11/2022	15/11/2022
					CT3001	Implantable metal	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	15/11/2022	15/11/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	15/11/2022	15/11/2022
					CT3002	Implantable synthetic polymer	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	15/11/2022	15/11/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	15/11/2022	15/11/2022
					CT3006	Implantable inorganic material	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	15/11/2022	15/11/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	15/11/2022	15/11/2022
					CT314	Electrically powered	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	15/11/2022	15/11/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	15/11/2022	15/11/2022
					CT321	Long-term surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	15/11/2022	15/11/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	15/11/2022	15/11/2022
					CT449	Generators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	15/11/2022	15/11/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	15/11/2022	15/11/2022
					CT981	Single use	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	15/11/2022	15/11/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	15/11/2022	15/11/2022
					CT983	Surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	15/11/2022	15/11/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	15/11/2022	15/11/2022
11626	2022		47265	Dual-chamber implantable pacemaker, rate-responsive	CT141	Electrodes/Leads and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	05/11/2022	05/11/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	05/11/2022	05/11/2022
					CT145	Cardiovascular devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	05/11/2022	05/11/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	05/11/2022	05/11/2022
					CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	05/11/2022	05/11/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	05/11/2022	05/11/2022
					CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	05/11/2022	05/11/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	05/11/2022	05/11/2022
					CT233	Surgical	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	05/11/2022	05/11/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	05/11/2022	05/11/2022
					CT2406	Implantable	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	05/11/2022	05/11/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	05/11/2022	05/11/2022
					CT2407	Active implantable	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	05/11/2022	05/11/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	05/11/2022	05/11/2022
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	05/11/2022	05/11/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	05/11/2022	05/11/2022
					CT2727	Active implant devices/ancillaries	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	05/11/2022	05/11/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	05/11/2022	05/11/2022
					CT2750	Implantable cardiac electrical stimulators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	05/11/2022	05/11/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	05/11/2022	05/11/2022
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	05/11/2022	05/11/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	05/11/2022	05/11/2022
					CT2985	Active	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	05/11/2022	05/11/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	05/11/2022	05/11/2022
					CT3001	Implantable metal	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	05/11/2022	05/11/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	05/11/2022	05/11/2022
					CT3002	Implantable synthetic polymer	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	05/11/2022	05/11/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	05/11/2022	05/11/2022
					CT3006	Implantable inorganic material	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	05/11/2022	05/11/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	05/11/2022	05/11/2022
					CT314	Electrically powered	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	05/11/2022	05/11/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	05/11/2022	05/11/2022
					CT321	Long-term surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	05/11/2022	05/11/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	05/11/2022	05/11/2022
					CT449	Generators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	05/11/2022	05/11/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	05/11/2022	05/11/2022
					CT981	Single use	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	05/11/2022	05/11/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	05/11/2022	05/11/2022
					CT983	Surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	05/11/2022	05/11/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	05/11/2022	05/11/2022
11627	2022	All other reportable incidents	35039	Epicardial pacing lead	CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070909	Oversensing		

								A072201	High Impedance		
								A090801	Signal Artifact/Noise		
						F26	No Health Consequences or Impact	A070909	Oversensing		
								A072201	High Impedance		
								A090801	Signal Artifact/Noise		
11628	2022	All other reportable incidents	47267	Single-chamber implantable pacemaker, rate-responsive	CT141	Electrodes/Leads and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	01/11/2022	01/11/2022
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	01/11/2022	01/11/2022
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	01/11/2022	01/11/2022
					CT145	Cardiovascular devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	01/11/2022	01/11/2022
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	01/11/2022	01/11/2022
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	01/11/2022	01/11/2022
					CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	01/11/2022	01/11/2022
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	01/11/2022	01/11/2022
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	01/11/2022	01/11/2022
					CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	01/11/2022	01/11/2022
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	01/11/2022	01/11/2022
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	01/11/2022	01/11/2022
					CT233	Surgical	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	01/11/2022	01/11/2022
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	01/11/2022	01/11/2022
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	01/11/2022	01/11/2022
					CT2406	Implantable	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	01/11/2022	01/11/2022
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	01/11/2022	01/11/2022
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	01/11/2022	01/11/2022
					CT2407	Active implantable	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	01/11/2022	01/11/2022
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	01/11/2022	01/11/2022
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	01/11/2022	01/11/2022
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	01/11/2022	01/11/2022
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	01/11/2022	01/11/2022
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	01/11/2022	01/11/2022
					CT2727	Active implant devices/ancillaries	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	01/11/2022	01/11/2022
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	01/11/2022	01/11/2022
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	01/11/2022	01/11/2022
					CT2750	Implantable cardiac electrical stimulators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	01/11/2022	01/11/2022
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	01/11/2022	01/11/2022
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	01/11/2022	01/11/2022
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	01/11/2022	01/11/2022
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	01/11/2022	01/11/2022
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	01/11/2022	01/11/2022
					CT2985	Active	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	01/11/2022	01/11/2022
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	01/11/2022	01/11/2022
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	01/11/2022	01/11/2022
					CT3001	Implantable metal	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	01/11/2022	01/11/2022
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	01/11/2022	01/11/2022
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	01/11/2022	01/11/2022
					CT3002	Implantable synthetic polymer	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	01/11/2022	01/11/2022
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	01/11/2022	01/11/2022
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	01/11/2022	01/11/2022
					CT3006	Implantable inorganic material	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	01/11/2022	01/11/2022
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	01/11/2022	01/11/2022
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	01/11/2022	01/11/2022
					CT314	Electrically powered	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	01/11/2022	01/11/2022
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	01/11/2022	01/11/2022
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	01/11/2022	01/11/2022
					CT321	Long-term surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	01/11/2022	01/11/2022
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	01/11/2022	01/11/2022
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	01/11/2022	01/11/2022
					CT449	Generators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	01/11/2022	01/11/2022
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	01/11/2022	01/11/2022
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	01/11/2022	01/11/2022
					CT981	Single use	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	01/11/2022	01/11/2022
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	01/11/2022	01/11/2022
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	01/11/2022	01/11/2022
					CT983	Surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	01/11/2022	01/11/2022
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	01/11/2022	01/11/2022
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	01/11/2022	01/11/2022
11629	2022	All other reportable incidents	35852	Single-chamber implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem	15/11/2022	15/11/2022
									A070504	Premature Discharge of Battery	15/11/2022	15/11/2022
							F08	Hospitalization or Prolonged Hospitalization	A0705	Battery Problem	15/11/2022	15/11/2022
									A070504	Premature Discharge of Battery	15/11/2022	15/11/2022
							F1905	Device Revision or Replacement	A0705	Battery Problem	15/11/2022	15/11/2022
									A070504	Premature Discharge of Battery	15/11/2022	15/11/2022
11630	2022	All other reportable incidents	36241	Pacemaker/defibrillator lead	CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A072201	High Impedance		
							F26	No Health Consequences or Impact	A072201	High Impedance		
11631	2022	All other reportable incidents	47265	Dual-chamber implantable pacemaker, rate-responsive	CT141	Electrodes/Leads and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070101	Failure to Capture	08/11/2022	08/11/2022
									A0709	Device Sensing Problem	08/11/2022	08/11/2022
									A0712	Pacing Problem	08/11/2022	08/11/2022
							F26	No Health Consequences or Impact	A070101	Failure to Capture	08/11/2022	08/11/2022
									A0709	Device Sensing Problem	08/11/2022	08/11/2022
									A0712	Pacing Problem	08/11/2022	08/11/2022
					CT145	Cardiovascular devices	E2403	No Clinical Signs, Symptoms or Conditions	A070101	Failure to Capture	08/11/2022	08/11/2022
									A0709	Device Sensing Problem	08/11/2022	08/11/2022
									A0712	Pacing Problem	08/11/2022	08/11/2022
							F26	No Health Consequences or Impact	A070101	Failure to Capture	08/11/2022	08/11/2022
									A0709	Device Sensing Problem	08/11/2022	08/11/2022
									A0712	Pacing Problem	08/11/2022	08/11/2022
					CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070101	Failure to Capture	08/11/2022	08/11/2022
									A0709	Device Sensing Problem	08/11/2022	08/11/2022
									A0712	Pacing Problem	08/11/2022	08/11/2022

		F26	No Health Consequences or Impact	A070101	Failure to Capture	08/11/2022	08/11/2022
				A0709	Device Sensing Problem	08/11/2022	08/11/2022
				A0712	Pacing Problem	08/11/2022	08/11/2022
CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A070101	Failure to Capture	08/11/2022	08/11/2022
				A0709	Device Sensing Problem	08/11/2022	08/11/2022
				A0712	Pacing Problem	08/11/2022	08/11/2022
		F26	No Health Consequences or Impact	A070101	Failure to Capture	08/11/2022	08/11/2022
				A0709	Device Sensing Problem	08/11/2022	08/11/2022
				A0712	Pacing Problem	08/11/2022	08/11/2022
CT233	Surgical	E2403	No Clinical Signs, Symptoms or Conditions	A070101	Failure to Capture	08/11/2022	08/11/2022
				A0709	Device Sensing Problem	08/11/2022	08/11/2022
				A0712	Pacing Problem	08/11/2022	08/11/2022
		F26	No Health Consequences or Impact	A070101	Failure to Capture	08/11/2022	08/11/2022
				A0709	Device Sensing Problem	08/11/2022	08/11/2022
				A0712	Pacing Problem	08/11/2022	08/11/2022
CT2406	Implantable	E2403	No Clinical Signs, Symptoms or Conditions	A070101	Failure to Capture	08/11/2022	08/11/2022
				A0709	Device Sensing Problem	08/11/2022	08/11/2022
				A0712	Pacing Problem	08/11/2022	08/11/2022
		F26	No Health Consequences or Impact	A070101	Failure to Capture	08/11/2022	08/11/2022
				A0709	Device Sensing Problem	08/11/2022	08/11/2022
				A0712	Pacing Problem	08/11/2022	08/11/2022
CT2407	Active implantable	E2403	No Clinical Signs, Symptoms or Conditions	A070101	Failure to Capture	08/11/2022	08/11/2022
				A0709	Device Sensing Problem	08/11/2022	08/11/2022
				A0712	Pacing Problem	08/11/2022	08/11/2022
		F26	No Health Consequences or Impact	A070101	Failure to Capture	08/11/2022	08/11/2022
				A0709	Device Sensing Problem	08/11/2022	08/11/2022
				A0712	Pacing Problem	08/11/2022	08/11/2022
CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A070101	Failure to Capture	08/11/2022	08/11/2022
				A0709	Device Sensing Problem	08/11/2022	08/11/2022
				A0712	Pacing Problem	08/11/2022	08/11/2022
		F26	No Health Consequences or Impact	A070101	Failure to Capture	08/11/2022	08/11/2022
				A0709	Device Sensing Problem	08/11/2022	08/11/2022
				A0712	Pacing Problem	08/11/2022	08/11/2022
CT2727	Active implant devices/ancillaries	E2403	No Clinical Signs, Symptoms or Conditions	A070101	Failure to Capture	08/11/2022	08/11/2022
				A0709	Device Sensing Problem	08/11/2022	08/11/2022
				A0712	Pacing Problem	08/11/2022	08/11/2022
		F26	No Health Consequences or Impact	A070101	Failure to Capture	08/11/2022	08/11/2022
				A0709	Device Sensing Problem	08/11/2022	08/11/2022
				A0712	Pacing Problem	08/11/2022	08/11/2022
CT2750	Implantable cardiac electrical stimulators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070101	Failure to Capture	08/11/2022	08/11/2022
				A0709	Device Sensing Problem	08/11/2022	08/11/2022
				A0712	Pacing Problem	08/11/2022	08/11/2022
		F26	No Health Consequences or Impact	A070101	Failure to Capture	08/11/2022	08/11/2022
				A0709	Device Sensing Problem	08/11/2022	08/11/2022
				A0712	Pacing Problem	08/11/2022	08/11/2022
CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A070101	Failure to Capture	08/11/2022	08/11/2022
				A0709	Device Sensing Problem	08/11/2022	08/11/2022
				A0712	Pacing Problem	08/11/2022	08/11/2022
		F26	No Health Consequences or Impact	A070101	Failure to Capture	08/11/2022	08/11/2022
				A0709	Device Sensing Problem	08/11/2022	08/11/2022
				A0712	Pacing Problem	08/11/2022	08/11/2022
CT2985	Active	E2403	No Clinical Signs, Symptoms or Conditions	A070101	Failure to Capture	08/11/2022	08/11/2022
				A0709	Device Sensing Problem	08/11/2022	08/11/2022
				A0712	Pacing Problem	08/11/2022	08/11/2022
		F26	No Health Consequences or Impact	A070101	Failure to Capture	08/11/2022	08/11/2022
				A0709	Device Sensing Problem	08/11/2022	08/11/2022
				A0712	Pacing Problem	08/11/2022	08/11/2022
CT3001	Implantable metal	E2403	No Clinical Signs, Symptoms or Conditions	A070101	Failure to Capture	08/11/2022	08/11/2022
				A0709	Device Sensing Problem	08/11/2022	08/11/2022
				A0712	Pacing Problem	08/11/2022	08/11/2022
		F26	No Health Consequences or Impact	A070101	Failure to Capture	08/11/2022	08/11/2022
				A0709	Device Sensing Problem	08/11/2022	08/11/2022
				A0712	Pacing Problem	08/11/2022	08/11/2022
CT3002	Implantable synthetic polymer	E2403	No Clinical Signs, Symptoms or Conditions	A070101	Failure to Capture	08/11/2022	08/11/2022
				A0709	Device Sensing Problem	08/11/2022	08/11/2022
				A0712	Pacing Problem	08/11/2022	08/11/2022
		F26	No Health Consequences or Impact	A070101	Failure to Capture	08/11/2022	08/11/2022
				A0709	Device Sensing Problem	08/11/2022	08/11/2022
				A0712	Pacing Problem	08/11/2022	08/11/2022
CT3006	Implantable inorganic material	E2403	No Clinical Signs, Symptoms or Conditions	A070101	Failure to Capture	08/11/2022	08/11/2022
				A0709	Device Sensing Problem	08/11/2022	08/11/2022
				A0712	Pacing Problem	08/11/2022	08/11/2022
		F26	No Health Consequences or Impact	A070101	Failure to Capture	08/11/2022	08/11/2022
				A0709	Device Sensing Problem	08/11/2022	08/11/2022
				A0712	Pacing Problem	08/11/2022	08/11/2022
CT314	Electrically powered	E2403	No Clinical Signs, Symptoms or Conditions	A070101	Failure to Capture	08/11/2022	08/11/2022
				A0709	Device Sensing Problem	08/11/2022	08/11/2022
				A0712	Pacing Problem	08/11/2022	08/11/2022
		F26	No Health Consequences or Impact	A070101	Failure to Capture	08/11/2022	08/11/2022
				A0709	Device Sensing Problem	08/11/2022	08/11/2022
				A0712	Pacing Problem	08/11/2022	08/11/2022
CT321	Long-term surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A070101	Failure to Capture	08/11/2022	08/11/2022
				A0709	Device Sensing Problem	08/11/2022	08/11/2022
				A0712	Pacing Problem	08/11/2022	08/11/2022
		F26	No Health Consequences or Impact	A070101	Failure to Capture	08/11/2022	08/11/2022
				A0709	Device Sensing Problem	08/11/2022	08/11/2022
				A0712	Pacing Problem	08/11/2022	08/11/2022
CT449	Generators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070101	Failure to Capture	08/11/2022	08/11/2022

									A0709	Device Sensing Problem	08/11/2022	08/11/2022
									A0712	Pacing Problem	08/11/2022	08/11/2022
							F26	No Health Consequences or Impact	A070101	Failure to Capture	08/11/2022	08/11/2022
									A0709	Device Sensing Problem	08/11/2022	08/11/2022
									A0712	Pacing Problem	08/11/2022	08/11/2022
					CT981	Single use	E2403	No Clinical Signs, Symptoms or Conditions	A070101	Failure to Capture	08/11/2022	08/11/2022
									A0709	Device Sensing Problem	08/11/2022	08/11/2022
									A0712	Pacing Problem	08/11/2022	08/11/2022
							F26	No Health Consequences or Impact	A070101	Failure to Capture	08/11/2022	08/11/2022
									A0709	Device Sensing Problem	08/11/2022	08/11/2022
									A0712	Pacing Problem	08/11/2022	08/11/2022
					CT983	Surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A070101	Failure to Capture	08/11/2022	08/11/2022
									A0709	Device Sensing Problem	08/11/2022	08/11/2022
									A0712	Pacing Problem	08/11/2022	08/11/2022
							F26	No Health Consequences or Impact	A070101	Failure to Capture	08/11/2022	08/11/2022
									A0709	Device Sensing Problem	08/11/2022	08/11/2022
									A0712	Pacing Problem	08/11/2022	08/11/2022
11632	2022	All other reportable incidents	35223	Endocardial/interventricular septal pacing lead	CT1684	Cardiac pacemakers and associated devices	E0112	Dizziness	A090801	Signal Artifact/Noise		
							F08	Hospitalization or Prolonged Hospitalization	A090801	Signal Artifact/Noise		
							F12	Serious Injury/ Illness/ Impairment	A090801	Signal Artifact/Noise		
							F23	Unexpected Medical Intervention	A090801	Signal Artifact/Noise		
11633	2022	All other reportable incidents	47265	Dual-chamber implantable pacemaker, rate-responsive	CT141	Electrodes/Leads and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
					CT145	Cardiovascular devices	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
					CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
					CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
					CT233	Surgical	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
					CT2406	Implantable	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
					CT2407	Active implantable	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
					CT2727	Active implant devices/ancillaries	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
					CT2750	Implantable cardiac electrical stimulators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
					CT2985	Active	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
					CT3001	Implantable metal	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
					CT3002	Implantable synthetic polymer	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
					CT3006	Implantable inorganic material	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
					CT314	Electrically powered	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
					CT321	Long-term surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A0705	Battery Problem		

								A070504	Premature Discharge of Battery			
				CT449	Generators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem			
								A070504	Premature Discharge of Battery			
						F26	No Health Consequences or Impact	A0705	Battery Problem			
								A070504	Premature Discharge of Battery			
				CT981	Single use	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem			
								A070504	Premature Discharge of Battery			
						F26	No Health Consequences or Impact	A0705	Battery Problem			
								A070504	Premature Discharge of Battery			
				CT983	Surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem			
								A070504	Premature Discharge of Battery			
						F26	No Health Consequences or Impact	A0705	Battery Problem			
								A070504	Premature Discharge of Battery			
11634	2022	All other reportable incidents	35852	Single-chamber implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E2104	Electric Shock	A070909	Oversensing		
									A071302	Inappropriate/Inadequate Shock/Stimulation		
									A090801	Signal Artifact/Noise		
						F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A070909	Oversensing			
								A071302	Inappropriate/Inadequate Shock/Stimulation			
								A090801	Signal Artifact/Noise			
						F2203	Imaging Required	A070909	Oversensing			
								A071302	Inappropriate/Inadequate Shock/Stimulation			
								A090801	Signal Artifact/Noise			
						F23	Unexpected Medical Intervention	A070909	Oversensing			
								A071302	Inappropriate/Inadequate Shock/Stimulation			
								A090801	Signal Artifact/Noise			
11635	2022	All other reportable incidents	35852	Single-chamber implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E2104	Electric Shock	A070909	Oversensing		
									A071302	Inappropriate/Inadequate Shock/Stimulation		
									A090801	Signal Artifact/Noise		
						F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A070909	Oversensing			
								A071302	Inappropriate/Inadequate Shock/Stimulation			
								A090801	Signal Artifact/Noise			
						F2203	Imaging Required	A070909	Oversensing			
								A071302	Inappropriate/Inadequate Shock/Stimulation			
								A090801	Signal Artifact/Noise			
						F23	Unexpected Medical Intervention	A070909	Oversensing			
								A071302	Inappropriate/Inadequate Shock/Stimulation			
								A090801	Signal Artifact/Noise			
11636	2022	All other reportable incidents	17882	Physiologic-monitoring defibrillation system	CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A0719	Unexpected Shutdown		
							F27	Problem Identified During Non-Clinical Procedure	A0719	Unexpected Shutdown		
11637	2022	All other reportable incidents	36241	Pacemaker/defibrillator lead	CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A072201	High Impedance		
							F2203	Imaging Required	A072201	High Impedance		
11638	2022		47270	Cardiac resynchronization therapy implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	17/11/2022	17/11/2022
							F1901	Additional Surgery	A070504	Premature Discharge of Battery	17/11/2022	17/11/2022
11639	2022		47265	Dual-chamber implantable pacemaker, rate-responsive	CT141	Electrodes/Leads and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	11/11/2022	11/11/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	11/11/2022	11/11/2022
					CT145	Cardiovascular devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	11/11/2022	11/11/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	11/11/2022	11/11/2022
					CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	11/11/2022	11/11/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	11/11/2022	11/11/2022
					CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	11/11/2022	11/11/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	11/11/2022	11/11/2022
					CT233	Surgical	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	11/11/2022	11/11/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	11/11/2022	11/11/2022
					CT2406	Implantable	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	11/11/2022	11/11/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	11/11/2022	11/11/2022
					CT2407	Active implantable	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	11/11/2022	11/11/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	11/11/2022	11/11/2022
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	11/11/2022	11/11/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	11/11/2022	11/11/2022
					CT2727	Active implant devices/ancillaries	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	11/11/2022	11/11/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	11/11/2022	11/11/2022
					CT2750	Implantable cardiac electrical stimulators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	11/11/2022	11/11/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	11/11/2022	11/11/2022
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	11/11/2022	11/11/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	11/11/2022	11/11/2022
					CT2985	Active	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	11/11/2022	11/11/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	11/11/2022	11/11/2022
					CT3001	Implantable metal	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	11/11/2022	11/11/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	11/11/2022	11/11/2022
					CT3002	Implantable synthetic polymer	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	11/11/2022	11/11/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	11/11/2022	11/11/2022
					CT3006	Implantable inorganic material	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	11/11/2022	11/11/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	11/11/2022	11/11/2022
					CT314	Electrically powered	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	11/11/2022	11/11/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	11/11/2022	11/11/2022
					CT321	Long-term surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	11/11/2022	11/11/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	11/11/2022	11/11/2022
					CT449	Generators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	11/11/2022	11/11/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	11/11/2022	11/11/2022
					CT981	Single use	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	11/11/2022	11/11/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	11/11/2022	11/11/2022
					CT983	Surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	11/11/2022	11/11/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	11/11/2022	11/11/2022
11640	2022	All other reportable incidents	47265	Dual-chamber implantable pacemaker, rate-responsive	CT141	Electrodes/Leads and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	12/11/2022	12/11/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	12/11/2022	12/11/2022
					CT145	Cardiovascular devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	12/11/2022	12/11/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	12/11/2022	12/11/2022
					CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	12/11/2022	12/11/2022

							F1901	Additional Surgery	A24	No Apparent Device Problem	12/11/2022	12/11/2022
					CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	12/11/2022	12/11/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	12/11/2022	12/11/2022
					CT233	Surgical	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	12/11/2022	12/11/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	12/11/2022	12/11/2022
					CT2406	Implantable	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	12/11/2022	12/11/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	12/11/2022	12/11/2022
					CT2407	Active implantable	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	12/11/2022	12/11/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	12/11/2022	12/11/2022
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	12/11/2022	12/11/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	12/11/2022	12/11/2022
					CT2727	Active implant devices/ancillaries	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	12/11/2022	12/11/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	12/11/2022	12/11/2022
					CT2750	Implantable cardiac electrical stimulators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	12/11/2022	12/11/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	12/11/2022	12/11/2022
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	12/11/2022	12/11/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	12/11/2022	12/11/2022
					CT2985	Active	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	12/11/2022	12/11/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	12/11/2022	12/11/2022
					CT3001	Implantable metal	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	12/11/2022	12/11/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	12/11/2022	12/11/2022
					CT3002	Implantable synthetic polymer	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	12/11/2022	12/11/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	12/11/2022	12/11/2022
					CT3006	Implantable inorganic material	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	12/11/2022	12/11/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	12/11/2022	12/11/2022
					CT314	Electrically powered	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	12/11/2022	12/11/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	12/11/2022	12/11/2022
					CT321	Long-term surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	12/11/2022	12/11/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	12/11/2022	12/11/2022
					CT449	Generators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	12/11/2022	12/11/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	12/11/2022	12/11/2022
					CT981	Single use	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	12/11/2022	12/11/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	12/11/2022	12/11/2022
					CT983	Surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	12/11/2022	12/11/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	12/11/2022	12/11/2022
11641	2022		47265	Dual-chamber implantable pacemaker, rate-responsive	CT141	Electrodes/Leads and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	12/11/2022	12/11/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	12/11/2022	12/11/2022
					CT145	Cardiovascular devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	12/11/2022	12/11/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	12/11/2022	12/11/2022
					CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	12/11/2022	12/11/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	12/11/2022	12/11/2022
					CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	12/11/2022	12/11/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	12/11/2022	12/11/2022
					CT233	Surgical	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	12/11/2022	12/11/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	12/11/2022	12/11/2022
					CT2406	Implantable	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	12/11/2022	12/11/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	12/11/2022	12/11/2022
					CT2407	Active implantable	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	12/11/2022	12/11/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	12/11/2022	12/11/2022
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	12/11/2022	12/11/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	12/11/2022	12/11/2022
					CT2727	Active implant devices/ancillaries	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	12/11/2022	12/11/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	12/11/2022	12/11/2022
					CT2750	Implantable cardiac electrical stimulators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	12/11/2022	12/11/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	12/11/2022	12/11/2022
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	12/11/2022	12/11/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	12/11/2022	12/11/2022
					CT2985	Active	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	12/11/2022	12/11/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	12/11/2022	12/11/2022
					CT3001	Implantable metal	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	12/11/2022	12/11/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	12/11/2022	12/11/2022
					CT3002	Implantable synthetic polymer	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	12/11/2022	12/11/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	12/11/2022	12/11/2022
					CT3006	Implantable inorganic material	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	12/11/2022	12/11/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	12/11/2022	12/11/2022
					CT314	Electrically powered	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	12/11/2022	12/11/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	12/11/2022	12/11/2022
					CT321	Long-term surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	12/11/2022	12/11/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	12/11/2022	12/11/2022
					CT449	Generators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	12/11/2022	12/11/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	12/11/2022	12/11/2022
					CT981	Single use	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	12/11/2022	12/11/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	12/11/2022	12/11/2022
					CT983	Surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	12/11/2022	12/11/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	12/11/2022	12/11/2022
11642	2022	All other reportable incidents	17882	Physiologic-monitoring defibrillation system	CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A0722	Impedance Problem		
							F27	Problem Identified During Non-Clinical Procedure	A0722	Impedance Problem		
11643	2022	All other reportable incidents	37265	Dual-chamber implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E060109	Tachycardia	A0709	Device Sensing Problem		
									A071301	Failure to Deliver Shock/Stimulation		
							F1001	Absence of Treatment	A0709	Device Sensing Problem		
									A071301	Failure to Deliver Shock/Stimulation		
							F1203	Life Threatening Illness or Injury	A0709	Device Sensing Problem		
									A071301	Failure to Deliver Shock/Stimulation		
11644	2022	All other reportable incidents	35852	Single-chamber implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A070910	Undersensing		
							F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A070910	Undersensing		
11645	2022	All other reportable incidents	35852	Single-chamber implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A070910	Undersensing		
							F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A070910	Undersensing		
11646	2022	All other reportable incidents	35852	Single-chamber implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A1603	Failure of Device to Self-Test		

							F26	No Health Consequences or Impact	A1603	Failure of Device to Self-Test		
11647	2022	All other reportable incidents	47263	Cardiac resynchronization therapy implantable pacemaker	CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A0908	Incorrect, Inadequate or Imprecise Result or Readings		
							F08	Hospitalization or Prolonged Hospitalization	A0908	Incorrect, Inadequate or Imprecise Result or Readings		
					CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A0908	Incorrect, Inadequate or Imprecise Result or Readings		
							F08	Hospitalization or Prolonged Hospitalization	A0908	Incorrect, Inadequate or Imprecise Result or Readings		
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A0908	Incorrect, Inadequate or Imprecise Result or Readings		
							F08	Hospitalization or Prolonged Hospitalization	A0908	Incorrect, Inadequate or Imprecise Result or Readings		
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A0908	Incorrect, Inadequate or Imprecise Result or Readings		
							F08	Hospitalization or Prolonged Hospitalization	A0908	Incorrect, Inadequate or Imprecise Result or Readings		
11648	2022	All other reportable incidents	47265	Dual-chamber implantable pacemaker, rate-responsive	CT141	Electrodes/Leads and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem		
							F26	No Health Consequences or Impact	A0705	Battery Problem		
					CT145	Cardiovascular devices	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem		
							F26	No Health Consequences or Impact	A0705	Battery Problem		
					CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem		
							F26	No Health Consequences or Impact	A0705	Battery Problem		
					CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem		
							F26	No Health Consequences or Impact	A0705	Battery Problem		
					CT233	Surgical	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem		
							F26	No Health Consequences or Impact	A0705	Battery Problem		
					CT2406	Implantable	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem		
							F26	No Health Consequences or Impact	A0705	Battery Problem		
					CT2407	Active implantable	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem		
							F26	No Health Consequences or Impact	A0705	Battery Problem		
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem		
							F26	No Health Consequences or Impact	A0705	Battery Problem		
					CT2727	Active implant devices/ancillaries	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem		
							F26	No Health Consequences or Impact	A0705	Battery Problem		
					CT2750	Implantable cardiac electrical stimulators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem		
							F26	No Health Consequences or Impact	A0705	Battery Problem		
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem		
							F26	No Health Consequences or Impact	A0705	Battery Problem		
					CT2985	Active	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem		
							F26	No Health Consequences or Impact	A0705	Battery Problem		
					CT3001	Implantable metal	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem		
							F26	No Health Consequences or Impact	A0705	Battery Problem		
					CT3002	Implantable synthetic polymer	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem		
							F26	No Health Consequences or Impact	A0705	Battery Problem		
					CT3006	Implantable inorganic material	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem		
							F26	No Health Consequences or Impact	A0705	Battery Problem		
					CT314	Electrically powered	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem		
							F26	No Health Consequences or Impact	A0705	Battery Problem		
					CT321	Long-term surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem		
							F26	No Health Consequences or Impact	A0705	Battery Problem		
					CT449	Generators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem		
							F26	No Health Consequences or Impact	A0705	Battery Problem		
					CT981	Single use	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem		
							F26	No Health Consequences or Impact	A0705	Battery Problem		
					CT983	Surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem		
							F26	No Health Consequences or Impact	A0705	Battery Problem		
11649	2022	All other reportable incidents	35223	Endocardial/interventricular septal pacing lead	CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A071208	Pacing Intermittently		
							F1905	Device Revision or Replacement	A071208	Pacing Intermittently		
11650	2022	All other reportable incidents	47265	Dual-chamber implantable pacemaker, rate-responsive	CT141	Electrodes/Leads and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	16/11/2022	16/11/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	16/11/2022	16/11/2022
					CT145	Cardiovascular devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	16/11/2022	16/11/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	16/11/2022	16/11/2022
					CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	16/11/2022	16/11/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	16/11/2022	16/11/2022
					CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	16/11/2022	16/11/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	16/11/2022	16/11/2022
					CT233	Surgical	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	16/11/2022	16/11/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	16/11/2022	16/11/2022
					CT2406	Implantable	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	16/11/2022	16/11/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	16/11/2022	16/11/2022
					CT2407	Active implantable	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	16/11/2022	16/11/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	16/11/2022	16/11/2022
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	16/11/2022	16/11/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	16/11/2022	16/11/2022
					CT2727	Active implant devices/ancillaries	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	16/11/2022	16/11/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	16/11/2022	16/11/2022
					CT2750	Implantable cardiac electrical stimulators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	16/11/2022	16/11/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	16/11/2022	16/11/2022
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	16/11/2022	16/11/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	16/11/2022	16/11/2022
					CT2985	Active	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	16/11/2022	16/11/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	16/11/2022	16/11/2022
					CT3001	Implantable metal	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	16/11/2022	16/11/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	16/11/2022	16/11/2022
					CT3002	Implantable synthetic polymer	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	16/11/2022	16/11/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	16/11/2022	16/11/2022
					CT3006	Implantable inorganic material	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	16/11/2022	16/11/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	16/11/2022	16/11/2022
					CT314	Electrically powered	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	16/11/2022	16/11/2022

							F1901	Additional Surgery	A24	No Apparent Device Problem	16/11/2022	16/11/2022
					CT321	Long-term surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	16/11/2022	16/11/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	16/11/2022	16/11/2022
					CT449	Generators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	16/11/2022	16/11/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	16/11/2022	16/11/2022
					CT981	Single use	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	16/11/2022	16/11/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	16/11/2022	16/11/2022
					CT983	Surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	16/11/2022	16/11/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	16/11/2022	16/11/2022
11651	2022	All other reportable incidents	35852	Single-chamber implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E2104	Electric Shock	A0709	Device Sensing Problem		
									A070909	Oversensing		
									A071302	Inappropriate/Inadequate Shock/Stimulation		
									A090801	Signal Artifact/Noise		
									A1301	Failure to Read Input Signal		
							F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A0709	Device Sensing Problem		
									A070909	Oversensing		
									A071302	Inappropriate/Inadequate Shock/Stimulation		
									A090801	Signal Artifact/Noise		
									A1301	Failure to Read Input Signal		
							F2203	Imaging Required	A0709	Device Sensing Problem		
									A070909	Oversensing		
									A071302	Inappropriate/Inadequate Shock/Stimulation		
									A090801	Signal Artifact/Noise		
									A1301	Failure to Read Input Signal		
							F23	Unexpected Medical Intervention	A0709	Device Sensing Problem		
									A070909	Oversensing		
									A071302	Inappropriate/Inadequate Shock/Stimulation		
									A090801	Signal Artifact/Noise		
									A1301	Failure to Read Input Signal		
11652	2022	All other reportable incidents	47265	Dual-chamber implantable pacemaker, rate-responsive	CT141	Electrodes/Leads and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
					CT145	Cardiovascular devices	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
					CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
					CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
					CT233	Surgical	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
					CT2406	Implantable	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
					CT2407	Active implantable	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
					CT2727	Active implant devices/ancillaries	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
					CT2750	Implantable cardiac electrical stimulators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
					CT2985	Active	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
					CT3001	Implantable metal	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
					CT3002	Implantable synthetic polymer	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
					CT3006	Implantable inorganic material	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A0705	Battery Problem		

									A070504	Premature Discharge of Battery		
					CT314	Electrically powered	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
					CT321	Long-term surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
					CT449	Generators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
					CT981	Single use	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
					CT983	Surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
11653	2022	All other reportable incidents	47265	Dual-chamber implantable pacemaker, rate-responsive	CT141	Electrodes/Leads and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	25/10/2022	25/10/2022
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	25/10/2022	25/10/2022
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	25/10/2022	25/10/2022
					CT145	Cardiovascular devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	25/10/2022	25/10/2022
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	25/10/2022	25/10/2022
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	25/10/2022	25/10/2022
					CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	25/10/2022	25/10/2022
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	25/10/2022	25/10/2022
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	25/10/2022	25/10/2022
					CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	25/10/2022	25/10/2022
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	25/10/2022	25/10/2022
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	25/10/2022	25/10/2022
					CT233	Surgical	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	25/10/2022	25/10/2022
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	25/10/2022	25/10/2022
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	25/10/2022	25/10/2022
					CT2406	Implantable	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	25/10/2022	25/10/2022
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	25/10/2022	25/10/2022
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	25/10/2022	25/10/2022
					CT2407	Active implantable	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	25/10/2022	25/10/2022
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	25/10/2022	25/10/2022
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	25/10/2022	25/10/2022
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	25/10/2022	25/10/2022
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	25/10/2022	25/10/2022
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	25/10/2022	25/10/2022
					CT2727	Active implant devices/ancillaries	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	25/10/2022	25/10/2022
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	25/10/2022	25/10/2022
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	25/10/2022	25/10/2022
					CT2750	Implantable cardiac electrical stimulators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	25/10/2022	25/10/2022
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	25/10/2022	25/10/2022
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	25/10/2022	25/10/2022
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	25/10/2022	25/10/2022
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	25/10/2022	25/10/2022
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	25/10/2022	25/10/2022
					CT2985	Active	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	25/10/2022	25/10/2022
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	25/10/2022	25/10/2022
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	25/10/2022	25/10/2022
					CT3001	Implantable metal	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	25/10/2022	25/10/2022
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	25/10/2022	25/10/2022
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	25/10/2022	25/10/2022
					CT3002	Implantable synthetic polymer	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	25/10/2022	25/10/2022
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	25/10/2022	25/10/2022
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	25/10/2022	25/10/2022
					CT3006	Implantable inorganic material	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	25/10/2022	25/10/2022
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	25/10/2022	25/10/2022
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	25/10/2022	25/10/2022
					CT314	Electrically powered	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	25/10/2022	25/10/2022
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	25/10/2022	25/10/2022
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	25/10/2022	25/10/2022
					CT321	Long-term surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	25/10/2022	25/10/2022
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	25/10/2022	25/10/2022
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	25/10/2022	25/10/2022
					CT449	Generators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	25/10/2022	25/10/2022
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	25/10/2022	25/10/2022
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	25/10/2022	25/10/2022
					CT981	Single use	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	25/10/2022	25/10/2022
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	25/10/2022	25/10/2022
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	25/10/2022	25/10/2022
					CT983	Surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	25/10/2022	25/10/2022
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	25/10/2022	25/10/2022
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	25/10/2022	25/10/2022
11654	2022	All other reportable incidents	35852	Single-chamber implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
									A1304	Telemetry Discrepancy		
							F26	No Health Consequences or Impact	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
									A1304	Telemetry Discrepancy		
11655	2022	All other reportable incidents	35223	Endocardial/interventricular septal pacing lead	CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A072202	Low Impedance		

									A090801	Signal Artifact/Noise		
							F26	No Health Consequences or Impact	A072202	Low Impedance		
									A090801	Signal Artifact/Noise		
11656	2022	All other reportable incidents	47265	Dual-chamber implantable pacemaker, rate-responsive	CT141	Electrodes/Leads and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	25/08/2022	25/08/2022
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	25/08/2022	25/08/2022
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	25/08/2022	25/08/2022
					CT145	Cardiovascular devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	25/08/2022	25/08/2022
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	25/08/2022	25/08/2022
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	25/08/2022	25/08/2022
					CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	25/08/2022	25/08/2022
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	25/08/2022	25/08/2022
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	25/08/2022	25/08/2022
					CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	25/08/2022	25/08/2022
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	25/08/2022	25/08/2022
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	25/08/2022	25/08/2022
					CT233	Surgical	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	25/08/2022	25/08/2022
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	25/08/2022	25/08/2022
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	25/08/2022	25/08/2022
					CT2406	Implantable	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	25/08/2022	25/08/2022
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	25/08/2022	25/08/2022
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	25/08/2022	25/08/2022
					CT2407	Active implantable	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	25/08/2022	25/08/2022
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	25/08/2022	25/08/2022
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	25/08/2022	25/08/2022
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	25/08/2022	25/08/2022
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	25/08/2022	25/08/2022
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	25/08/2022	25/08/2022
					CT2727	Active implant devices/ancillaries	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	25/08/2022	25/08/2022
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	25/08/2022	25/08/2022
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	25/08/2022	25/08/2022
					CT2750	Implantable cardiac electrical stimulators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	25/08/2022	25/08/2022
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	25/08/2022	25/08/2022
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	25/08/2022	25/08/2022
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	25/08/2022	25/08/2022
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	25/08/2022	25/08/2022
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	25/08/2022	25/08/2022
					CT2985	Active	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	25/08/2022	25/08/2022
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	25/08/2022	25/08/2022
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	25/08/2022	25/08/2022
					CT3001	Implantable metal	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	25/08/2022	25/08/2022
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	25/08/2022	25/08/2022
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	25/08/2022	25/08/2022
					CT3002	Implantable synthetic polymer	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	25/08/2022	25/08/2022
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	25/08/2022	25/08/2022
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	25/08/2022	25/08/2022
					CT3006	Implantable inorganic material	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	25/08/2022	25/08/2022
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	25/08/2022	25/08/2022
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	25/08/2022	25/08/2022
					CT314	Electrically powered	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	25/08/2022	25/08/2022
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	25/08/2022	25/08/2022
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	25/08/2022	25/08/2022
					CT321	Long-term surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	25/08/2022	25/08/2022
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	25/08/2022	25/08/2022
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	25/08/2022	25/08/2022
					CT449	Generators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	25/08/2022	25/08/2022
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	25/08/2022	25/08/2022
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	25/08/2022	25/08/2022
					CT981	Single use	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	25/08/2022	25/08/2022
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	25/08/2022	25/08/2022
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	25/08/2022	25/08/2022
					CT983	Surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	25/08/2022	25/08/2022
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	25/08/2022	25/08/2022
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	25/08/2022	25/08/2022
11657	2022	All other reportable incidents	47265	Dual-chamber implantable pacemaker, rate-responsive	CT141	Electrodes/Leads and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	22/11/2022	22/11/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	22/11/2022	22/11/2022
					CT145	Cardiovascular devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	22/11/2022	22/11/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	22/11/2022	22/11/2022
					CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	22/11/2022	22/11/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	22/11/2022	22/11/2022
					CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	22/11/2022	22/11/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	22/11/2022	22/11/2022
					CT233	Surgical	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	22/11/2022	22/11/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	22/11/2022	22/11/2022
					CT2406	Implantable	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	22/11/2022	22/11/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	22/11/2022	22/11/2022
					CT2407	Active implantable	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	22/11/2022	22/11/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	22/11/2022	22/11/2022
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	22/11/2022	22/11/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	22/11/2022	22/11/2022
					CT2727	Active implant devices/ancillaries	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	22/11/2022	22/11/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	22/11/2022	22/11/2022
					CT2750	Implantable cardiac electrical stimulators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	22/11/2022	22/11/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	22/11/2022	22/11/2022
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	22/11/2022	22/11/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	22/11/2022	22/11/2022
					CT2985	Active	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	22/11/2022	22/11/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	22/11/2022	22/11/2022
					CT3001	Implantable metal	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	22/11/2022	22/11/2022

							F1901	Additional Surgery	A24	No Apparent Device Problem	22/11/2022	22/11/2022
					CT3002	Implantable synthetic polymer	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	22/11/2022	22/11/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	22/11/2022	22/11/2022
					CT3006	Implantable inorganic material	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	22/11/2022	22/11/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	22/11/2022	22/11/2022
					CT314	Electrically powered	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	22/11/2022	22/11/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	22/11/2022	22/11/2022
					CT321	Long-term surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	22/11/2022	22/11/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	22/11/2022	22/11/2022
					CT449	Generators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	22/11/2022	22/11/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	22/11/2022	22/11/2022
					CT981	Single use	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	22/11/2022	22/11/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	22/11/2022	22/11/2022
					CT983	Surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	22/11/2022	22/11/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	22/11/2022	22/11/2022
11658	2022	Unanticipated serious deterioration in state of health	35852	Single-chamber implantable defibrillator	CT2866	Implantable cardiac electrical stimulators						
11659	2022	Unanticipated serious deterioration in state of health			CT257	Implantable cardiac pacemakers			A26	Insufficient Device Problem Information		
11660	2022	All other reportable incidents	47267	Single-chamber implantable pacemaker, rate-responsive	CT141	Electrodes/Leads and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A07	Electrical /Electronic Property Problem	27/10/2022	27/10/2022
									A0705	Battery Problem	27/10/2022	27/10/2022
									A070504	Premature Discharge of Battery	27/10/2022	27/10/2022
							F08	Hospitalization or Prolonged Hospitalization	A07	Electrical /Electronic Property Problem	27/10/2022	27/10/2022
									A0705	Battery Problem	27/10/2022	27/10/2022
									A070504	Premature Discharge of Battery	27/10/2022	27/10/2022
							F1905	Device Revision or Replacement	A07	Electrical /Electronic Property Problem	27/10/2022	27/10/2022
									A0705	Battery Problem	27/10/2022	27/10/2022
									A070504	Premature Discharge of Battery	27/10/2022	27/10/2022
					CT145	Cardiovascular devices	E2403	No Clinical Signs, Symptoms or Conditions	A07	Electrical /Electronic Property Problem	27/10/2022	27/10/2022
									A0705	Battery Problem	27/10/2022	27/10/2022
									A070504	Premature Discharge of Battery	27/10/2022	27/10/2022
							F08	Hospitalization or Prolonged Hospitalization	A07	Electrical /Electronic Property Problem	27/10/2022	27/10/2022
									A0705	Battery Problem	27/10/2022	27/10/2022
									A070504	Premature Discharge of Battery	27/10/2022	27/10/2022
							F1905	Device Revision or Replacement	A07	Electrical /Electronic Property Problem	27/10/2022	27/10/2022
									A0705	Battery Problem	27/10/2022	27/10/2022
									A070504	Premature Discharge of Battery	27/10/2022	27/10/2022
					CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A07	Electrical /Electronic Property Problem	27/10/2022	27/10/2022
									A0705	Battery Problem	27/10/2022	27/10/2022
									A070504	Premature Discharge of Battery	27/10/2022	27/10/2022
							F08	Hospitalization or Prolonged Hospitalization	A07	Electrical /Electronic Property Problem	27/10/2022	27/10/2022
									A0705	Battery Problem	27/10/2022	27/10/2022
									A070504	Premature Discharge of Battery	27/10/2022	27/10/2022
							F1905	Device Revision or Replacement	A07	Electrical /Electronic Property Problem	27/10/2022	27/10/2022
									A0705	Battery Problem	27/10/2022	27/10/2022
									A070504	Premature Discharge of Battery	27/10/2022	27/10/2022
					CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A07	Electrical /Electronic Property Problem	27/10/2022	27/10/2022
									A0705	Battery Problem	27/10/2022	27/10/2022
									A070504	Premature Discharge of Battery	27/10/2022	27/10/2022
							F08	Hospitalization or Prolonged Hospitalization	A07	Electrical /Electronic Property Problem	27/10/2022	27/10/2022
									A0705	Battery Problem	27/10/2022	27/10/2022
									A070504	Premature Discharge of Battery	27/10/2022	27/10/2022
							F1905	Device Revision or Replacement	A07	Electrical /Electronic Property Problem	27/10/2022	27/10/2022
									A0705	Battery Problem	27/10/2022	27/10/2022
									A070504	Premature Discharge of Battery	27/10/2022	27/10/2022
					CT233	Surgical	E2403	No Clinical Signs, Symptoms or Conditions	A07	Electrical /Electronic Property Problem	27/10/2022	27/10/2022
									A0705	Battery Problem	27/10/2022	27/10/2022
									A070504	Premature Discharge of Battery	27/10/2022	27/10/2022
							F08	Hospitalization or Prolonged Hospitalization	A07	Electrical /Electronic Property Problem	27/10/2022	27/10/2022
									A0705	Battery Problem	27/10/2022	27/10/2022
									A070504	Premature Discharge of Battery	27/10/2022	27/10/2022
							F1905	Device Revision or Replacement	A07	Electrical /Electronic Property Problem	27/10/2022	27/10/2022
									A0705	Battery Problem	27/10/2022	27/10/2022
									A070504	Premature Discharge of Battery	27/10/2022	27/10/2022
					CT2406	Implantable	E2403	No Clinical Signs, Symptoms or Conditions	A07	Electrical /Electronic Property Problem	27/10/2022	27/10/2022
									A0705	Battery Problem	27/10/2022	27/10/2022
									A070504	Premature Discharge of Battery	27/10/2022	27/10/2022
							F08	Hospitalization or Prolonged Hospitalization	A07	Electrical /Electronic Property Problem	27/10/2022	27/10/2022
									A0705	Battery Problem	27/10/2022	27/10/2022
									A070504	Premature Discharge of Battery	27/10/2022	27/10/2022
							F1905	Device Revision or Replacement	A07	Electrical /Electronic Property Problem	27/10/2022	27/10/2022
									A0705	Battery Problem	27/10/2022	27/10/2022
									A070504	Premature Discharge of Battery	27/10/2022	27/10/2022
					CT2407	Active implantable	E2403	No Clinical Signs, Symptoms or Conditions	A07	Electrical /Electronic Property Problem	27/10/2022	27/10/2022
									A0705	Battery Problem	27/10/2022	27/10/2022
									A070504	Premature Discharge of Battery	27/10/2022	27/10/2022
							F08	Hospitalization or Prolonged Hospitalization	A07	Electrical /Electronic Property Problem	27/10/2022	27/10/2022
									A0705	Battery Problem	27/10/2022	27/10/2022
									A070504	Premature Discharge of Battery	27/10/2022	27/10/2022
							F1905	Device Revision or Replacement	A07	Electrical /Electronic Property Problem	27/10/2022	27/10/2022
									A0705	Battery Problem	27/10/2022	27/10/2022
									A070504	Premature Discharge of Battery	27/10/2022	27/10/2022
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A07	Electrical /Electronic Property Problem	27/10/2022	27/10/2022
									A0705	Battery Problem	27/10/2022	27/10/2022
									A070504	Premature Discharge of Battery	27/10/2022	27/10/2022
							F08	Hospitalization or Prolonged Hospitalization	A07	Electrical /Electronic Property Problem	27/10/2022	27/10/2022
									A0705	Battery Problem	27/10/2022	27/10/2022
									A070504	Premature Discharge of Battery	27/10/2022	27/10/2022

						F1905	Device Revision or Replacement	A07	Electrical /Electronic Property Problem	27/10/2022	27/10/2022
								A0705	Battery Problem	27/10/2022	27/10/2022
								A070504	Premature Discharge of Battery	27/10/2022	27/10/2022
					CT2727	Active implant devices/ancillaries	E2403	No Clinical Signs, Symptoms or Conditions	A07	Electrical /Electronic Property Problem	27/10/2022	27/10/2022
									A0705	Battery Problem	27/10/2022	27/10/2022
									A070504	Premature Discharge of Battery	27/10/2022	27/10/2022
							F08	Hospitalization or Prolonged Hospitalization	A07	Electrical /Electronic Property Problem	27/10/2022	27/10/2022
									A0705	Battery Problem	27/10/2022	27/10/2022
									A070504	Premature Discharge of Battery	27/10/2022	27/10/2022
							F1905	Device Revision or Replacement	A07	Electrical /Electronic Property Problem	27/10/2022	27/10/2022
									A0705	Battery Problem	27/10/2022	27/10/2022
									A070504	Premature Discharge of Battery	27/10/2022	27/10/2022
					CT2750	Implantable cardiac electrical stimulators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A07	Electrical /Electronic Property Problem	27/10/2022	27/10/2022
									A0705	Battery Problem	27/10/2022	27/10/2022
									A070504	Premature Discharge of Battery	27/10/2022	27/10/2022
							F08	Hospitalization or Prolonged Hospitalization	A07	Electrical /Electronic Property Problem	27/10/2022	27/10/2022
									A0705	Battery Problem	27/10/2022	27/10/2022
									A070504	Premature Discharge of Battery	27/10/2022	27/10/2022
							F1905	Device Revision or Replacement	A07	Electrical /Electronic Property Problem	27/10/2022	27/10/2022
									A0705	Battery Problem	27/10/2022	27/10/2022
									A070504	Premature Discharge of Battery	27/10/2022	27/10/2022
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A07	Electrical /Electronic Property Problem	27/10/2022	27/10/2022
									A0705	Battery Problem	27/10/2022	27/10/2022
									A070504	Premature Discharge of Battery	27/10/2022	27/10/2022
							F08	Hospitalization or Prolonged Hospitalization	A07	Electrical /Electronic Property Problem	27/10/2022	27/10/2022
									A0705	Battery Problem	27/10/2022	27/10/2022
									A070504	Premature Discharge of Battery	27/10/2022	27/10/2022
							F1905	Device Revision or Replacement	A07	Electrical /Electronic Property Problem	27/10/2022	27/10/2022
									A0705	Battery Problem	27/10/2022	27/10/2022
									A070504	Premature Discharge of Battery	27/10/2022	27/10/2022
					CT2985	Active	E2403	No Clinical Signs, Symptoms or Conditions	A07	Electrical /Electronic Property Problem	27/10/2022	27/10/2022
									A0705	Battery Problem	27/10/2022	27/10/2022
									A070504	Premature Discharge of Battery	27/10/2022	27/10/2022
							F08	Hospitalization or Prolonged Hospitalization	A07	Electrical /Electronic Property Problem	27/10/2022	27/10/2022
									A0705	Battery Problem	27/10/2022	27/10/2022
									A070504	Premature Discharge of Battery	27/10/2022	27/10/2022
							F1905	Device Revision or Replacement	A07	Electrical /Electronic Property Problem	27/10/2022	27/10/2022
									A0705	Battery Problem	27/10/2022	27/10/2022
									A070504	Premature Discharge of Battery	27/10/2022	27/10/2022
					CT3001	Implantable metal	E2403	No Clinical Signs, Symptoms or Conditions	A07	Electrical /Electronic Property Problem	27/10/2022	27/10/2022
									A0705	Battery Problem	27/10/2022	27/10/2022
									A070504	Premature Discharge of Battery	27/10/2022	27/10/2022
							F08	Hospitalization or Prolonged Hospitalization	A07	Electrical /Electronic Property Problem	27/10/2022	27/10/2022
									A0705	Battery Problem	27/10/2022	27/10/2022
									A070504	Premature Discharge of Battery	27/10/2022	27/10/2022
							F1905	Device Revision or Replacement	A07	Electrical /Electronic Property Problem	27/10/2022	27/10/2022
									A0705	Battery Problem	27/10/2022	27/10/2022
									A070504	Premature Discharge of Battery	27/10/2022	27/10/2022
					CT3002	Implantable synthetic polymer	E2403	No Clinical Signs, Symptoms or Conditions	A07	Electrical /Electronic Property Problem	27/10/2022	27/10/2022
									A0705	Battery Problem	27/10/2022	27/10/2022
									A070504	Premature Discharge of Battery	27/10/2022	27/10/2022
							F08	Hospitalization or Prolonged Hospitalization	A07	Electrical /Electronic Property Problem	27/10/2022	27/10/2022
									A0705	Battery Problem	27/10/2022	27/10/2022
									A070504	Premature Discharge of Battery	27/10/2022	27/10/2022
							F1905	Device Revision or Replacement	A07	Electrical /Electronic Property Problem	27/10/2022	27/10/2022
									A0705	Battery Problem	27/10/2022	27/10/2022
									A070504	Premature Discharge of Battery	27/10/2022	27/10/2022
					CT3006	Implantable inorganic material	E2403	No Clinical Signs, Symptoms or Conditions	A07	Electrical /Electronic Property Problem	27/10/2022	27/10/2022
									A0705	Battery Problem	27/10/2022	27/10/2022
									A070504	Premature Discharge of Battery	27/10/2022	27/10/2022
							F08	Hospitalization or Prolonged Hospitalization	A07	Electrical /Electronic Property Problem	27/10/2022	27/10/2022
									A0705	Battery Problem	27/10/2022	27/10/2022
									A070504	Premature Discharge of Battery	27/10/2022	27/10/2022
							F1905	Device Revision or Replacement	A07	Electrical /Electronic Property Problem	27/10/2022	27/10/2022
									A0705	Battery Problem	27/10/2022	27/10/2022
									A070504	Premature Discharge of Battery	27/10/2022	27/10/2022
					CT314	Electrically powered	E2403	No Clinical Signs, Symptoms or Conditions	A07	Electrical /Electronic Property Problem	27/10/2022	27/10/2022
									A0705	Battery Problem	27/10/2022	27/10/2022
									A070504	Premature Discharge of Battery	27/10/2022	27/10/2022
							F08	Hospitalization or Prolonged Hospitalization	A07	Electrical /Electronic Property Problem	27/10/2022	27/10/2022
									A0705	Battery Problem	27/10/2022	27/10/2022
									A070504	Premature Discharge of Battery	27/10/2022	27/10/2022
							F1905	Device Revision or Replacement	A07	Electrical /Electronic Property Problem	27/10/2022	27/10/2022
									A0705	Battery Problem	27/10/2022	27/10/2022
									A070504	Premature Discharge of Battery	27/10/2022	27/10/2022
					CT321	Long-term surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A07	Electrical /Electronic Property Problem	27/10/2022	27/10/2022
									A0705	Battery Problem	27/10/2022	27/10/2022
									A070504	Premature Discharge of Battery	27/10/2022	27/10/2022
							F08	Hospitalization or Prolonged Hospitalization	A07	Electrical /Electronic Property Problem	27/10/2022	27/10/2022
									A0705	Battery Problem	27/10/2022	27/10/2022
									A070504	Premature Discharge of Battery	27/10/2022	27/10/2022
							F1905	Device Revision or Replacement	A07	Electrical /Electronic Property Problem	27/10/2022	27/10/2022
									A0705	Battery Problem	27/10/2022	27/10/2022
									A070504	Premature Discharge of Battery	27/10/2022	27/10/2022
					CT449	Generators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A07	Electrical /Electronic Property Problem	27/10/2022	27/10/2022
									A0705	Battery Problem	27/10/2022	27/10/2022
									A070504	Premature Discharge of Battery	27/10/2022	27/10/2022
							F08	Hospitalization or Prolonged Hospitalization	A07	Electrical /Electronic Property Problem	27/10/2022	27/10/2022

									A0705	Battery Problem	27/10/2022	27/10/2022
									A070504	Premature Discharge of Battery	27/10/2022	27/10/2022
							F1905	Device Revision or Replacement	A07	Electrical /Electronic Property Problem	27/10/2022	27/10/2022
									A0705	Battery Problem	27/10/2022	27/10/2022
									A070504	Premature Discharge of Battery	27/10/2022	27/10/2022
					CT981	Single use	E2403	No Clinical Signs, Symptoms or Conditions	A07	Electrical /Electronic Property Problem	27/10/2022	27/10/2022
									A0705	Battery Problem	27/10/2022	27/10/2022
									A070504	Premature Discharge of Battery	27/10/2022	27/10/2022
							F08	Hospitalization or Prolonged Hospitalization	A07	Electrical /Electronic Property Problem	27/10/2022	27/10/2022
									A0705	Battery Problem	27/10/2022	27/10/2022
									A070504	Premature Discharge of Battery	27/10/2022	27/10/2022
							F1905	Device Revision or Replacement	A07	Electrical /Electronic Property Problem	27/10/2022	27/10/2022
									A0705	Battery Problem	27/10/2022	27/10/2022
									A070504	Premature Discharge of Battery	27/10/2022	27/10/2022
					CT983	Surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A07	Electrical /Electronic Property Problem	27/10/2022	27/10/2022
									A0705	Battery Problem	27/10/2022	27/10/2022
									A070504	Premature Discharge of Battery	27/10/2022	27/10/2022
							F08	Hospitalization or Prolonged Hospitalization	A07	Electrical /Electronic Property Problem	27/10/2022	27/10/2022
									A0705	Battery Problem	27/10/2022	27/10/2022
									A070504	Premature Discharge of Battery	27/10/2022	27/10/2022
							F1905	Device Revision or Replacement	A07	Electrical /Electronic Property Problem	27/10/2022	27/10/2022
									A0705	Battery Problem	27/10/2022	27/10/2022
									A070504	Premature Discharge of Battery	27/10/2022	27/10/2022
11661	2022	All other reportable incidents	47265	Dual-chamber implantable pacemaker, rate-responsive	CT141	Electrodes/Leads and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A07	Electrical /Electronic Property Problem		
									A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
							F24	Insufficient Health Impact Information	A07	Electrical /Electronic Property Problem		
									A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
					CT145	Cardiovascular devices	E2403	No Clinical Signs, Symptoms or Conditions	A07	Electrical /Electronic Property Problem		
									A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
							F24	Insufficient Health Impact Information	A07	Electrical /Electronic Property Problem		
									A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
					CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A07	Electrical /Electronic Property Problem		
									A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
							F24	Insufficient Health Impact Information	A07	Electrical /Electronic Property Problem		
									A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
					CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A07	Electrical /Electronic Property Problem		
									A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
							F24	Insufficient Health Impact Information	A07	Electrical /Electronic Property Problem		
									A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
					CT233	Surgical	E2403	No Clinical Signs, Symptoms or Conditions	A07	Electrical /Electronic Property Problem		
									A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
							F24	Insufficient Health Impact Information	A07	Electrical /Electronic Property Problem		
									A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
					CT2406	Implantable	E2403	No Clinical Signs, Symptoms or Conditions	A07	Electrical /Electronic Property Problem		
									A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
							F24	Insufficient Health Impact Information	A07	Electrical /Electronic Property Problem		
									A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
					CT2407	Active implantable	E2403	No Clinical Signs, Symptoms or Conditions	A07	Electrical /Electronic Property Problem		
									A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
							F24	Insufficient Health Impact Information	A07	Electrical /Electronic Property Problem		
									A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A07	Electrical /Electronic Property Problem		
									A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
							F24	Insufficient Health Impact Information	A07	Electrical /Electronic Property Problem		
									A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
					CT2727	Active implant devices/ancillaries	E2403	No Clinical Signs, Symptoms or Conditions	A07	Electrical /Electronic Property Problem		
									A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
							F24	Insufficient Health Impact Information	A07	Electrical /Electronic Property Problem		
									A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
					CT2750	Implantable cardiac electrical stimulators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A07	Electrical /Electronic Property Problem		
									A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
							F24	Insufficient Health Impact Information	A07	Electrical /Electronic Property Problem		
									A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A07	Electrical /Electronic Property Problem		
									A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
							F24	Insufficient Health Impact Information	A07	Electrical /Electronic Property Problem		
									A0705	Battery Problem		

									A070504	Premature Discharge of Battery		
					CT2985	Active	E2403	No Clinical Signs, Symptoms or Conditions	A07	Electrical /Electronic Property Problem		
									A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
							F24	Insufficient Health Impact Information	A07	Electrical /Electronic Property Problem		
									A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
					CT3001	Implantable metal	E2403	No Clinical Signs, Symptoms or Conditions	A07	Electrical /Electronic Property Problem		
									A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
							F24	Insufficient Health Impact Information	A07	Electrical /Electronic Property Problem		
									A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
					CT3002	Implantable synthetic polymer	E2403	No Clinical Signs, Symptoms or Conditions	A07	Electrical /Electronic Property Problem		
									A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
							F24	Insufficient Health Impact Information	A07	Electrical /Electronic Property Problem		
									A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
					CT3006	Implantable inorganic material	E2403	No Clinical Signs, Symptoms or Conditions	A07	Electrical /Electronic Property Problem		
									A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
							F24	Insufficient Health Impact Information	A07	Electrical /Electronic Property Problem		
									A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
					CT314	Electrically powered	E2403	No Clinical Signs, Symptoms or Conditions	A07	Electrical /Electronic Property Problem		
									A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
							F24	Insufficient Health Impact Information	A07	Electrical /Electronic Property Problem		
									A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
					CT321	Long-term surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A07	Electrical /Electronic Property Problem		
									A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
							F24	Insufficient Health Impact Information	A07	Electrical /Electronic Property Problem		
									A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
					CT449	Generators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A07	Electrical /Electronic Property Problem		
									A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
							F24	Insufficient Health Impact Information	A07	Electrical /Electronic Property Problem		
									A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
					CT981	Single use	E2403	No Clinical Signs, Symptoms or Conditions	A07	Electrical /Electronic Property Problem		
									A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
							F24	Insufficient Health Impact Information	A07	Electrical /Electronic Property Problem		
									A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
					CT983	Surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A07	Electrical /Electronic Property Problem		
									A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
							F24	Insufficient Health Impact Information	A07	Electrical /Electronic Property Problem		
									A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
11662	2022	All other reportable incidents	17882	Physiologic-monitoring defibrillation system	CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A11	Computer Software Problem		
							F27	Problem Identified During Non-Clinical Procedure	A11	Computer Software Problem		
11663	2022	All other reportable incidents	35223	Endocardial/interventricular septal pacing lead	CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A072202	Low Impedance		
							F12	Serious Injury/ Illness/ Impairment	A072202	Low Impedance		
							F23	Unexpected Medical Intervention	A072202	Low Impedance		
11664	2022	All other reportable incidents	47265	Dual-chamber implantable pacemaker, rate-responsive	CT141	Electrodes/Leads and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A0709	Device Sensing Problem	24/11/2022	24/11/2022
									A0712	Pacing Problem	24/11/2022	24/11/2022
							F08	Hospitalization or Prolonged Hospitalization	A0709	Device Sensing Problem	24/11/2022	24/11/2022
									A0712	Pacing Problem	24/11/2022	24/11/2022
							F19	Surgical Intervention	A0709	Device Sensing Problem	24/11/2022	24/11/2022
									A0712	Pacing Problem	24/11/2022	24/11/2022
					CT145	Cardiovascular devices	E2403	No Clinical Signs, Symptoms or Conditions	A0709	Device Sensing Problem	24/11/2022	24/11/2022
									A0712	Pacing Problem	24/11/2022	24/11/2022
							F08	Hospitalization or Prolonged Hospitalization	A0709	Device Sensing Problem	24/11/2022	24/11/2022
									A0712	Pacing Problem	24/11/2022	24/11/2022
							F19	Surgical Intervention	A0709	Device Sensing Problem	24/11/2022	24/11/2022
									A0712	Pacing Problem	24/11/2022	24/11/2022
					CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A0709	Device Sensing Problem	24/11/2022	24/11/2022
									A0712	Pacing Problem	24/11/2022	24/11/2022
							F08	Hospitalization or Prolonged Hospitalization	A0709	Device Sensing Problem	24/11/2022	24/11/2022
									A0712	Pacing Problem	24/11/2022	24/11/2022
							F19	Surgical Intervention	A0709	Device Sensing Problem	24/11/2022	24/11/2022
									A0712	Pacing Problem	24/11/2022	24/11/2022
					CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A0709	Device Sensing Problem	24/11/2022	24/11/2022
									A0712	Pacing Problem	24/11/2022	24/11/2022
							F08	Hospitalization or Prolonged Hospitalization	A0709	Device Sensing Problem	24/11/2022	24/11/2022
									A0712	Pacing Problem	24/11/2022	24/11/2022
							F19	Surgical Intervention	A0709	Device Sensing Problem	24/11/2022	24/11/2022
									A0712	Pacing Problem	24/11/2022	24/11/2022
					CT233	Surgical	E2403	No Clinical Signs, Symptoms or Conditions	A0709	Device Sensing Problem	24/11/2022	24/11/2022
									A0712	Pacing Problem	24/11/2022	24/11/2022
							F08	Hospitalization or Prolonged Hospitalization	A0709	Device Sensing Problem	24/11/2022	24/11/2022
									A0712	Pacing Problem	24/11/2022	24/11/2022

		F19	Surgical Intervention	A0709	Device Sensing Problem	24/11/2022	24/11/2022
				A0712	Pacing Problem	24/11/2022	24/11/2022
CT2406	Implantable	E2403	No Clinical Signs, Symptoms or Conditions	A0709	Device Sensing Problem	24/11/2022	24/11/2022
				A0712	Pacing Problem	24/11/2022	24/11/2022
		F08	Hospitalization or Prolonged Hospitalization	A0709	Device Sensing Problem	24/11/2022	24/11/2022
				A0712	Pacing Problem	24/11/2022	24/11/2022
		F19	Surgical Intervention	A0709	Device Sensing Problem	24/11/2022	24/11/2022
				A0712	Pacing Problem	24/11/2022	24/11/2022
CT2407	Active implantable	E2403	No Clinical Signs, Symptoms or Conditions	A0709	Device Sensing Problem	24/11/2022	24/11/2022
				A0712	Pacing Problem	24/11/2022	24/11/2022
		F08	Hospitalization or Prolonged Hospitalization	A0709	Device Sensing Problem	24/11/2022	24/11/2022
				A0712	Pacing Problem	24/11/2022	24/11/2022
		F19	Surgical Intervention	A0709	Device Sensing Problem	24/11/2022	24/11/2022
				A0712	Pacing Problem	24/11/2022	24/11/2022
CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A0709	Device Sensing Problem	24/11/2022	24/11/2022
				A0712	Pacing Problem	24/11/2022	24/11/2022
		F08	Hospitalization or Prolonged Hospitalization	A0709	Device Sensing Problem	24/11/2022	24/11/2022
				A0712	Pacing Problem	24/11/2022	24/11/2022
		F19	Surgical Intervention	A0709	Device Sensing Problem	24/11/2022	24/11/2022
				A0712	Pacing Problem	24/11/2022	24/11/2022
CT2727	Active implant devices/ancillaries	E2403	No Clinical Signs, Symptoms or Conditions	A0709	Device Sensing Problem	24/11/2022	24/11/2022
				A0712	Pacing Problem	24/11/2022	24/11/2022
		F08	Hospitalization or Prolonged Hospitalization	A0709	Device Sensing Problem	24/11/2022	24/11/2022
				A0712	Pacing Problem	24/11/2022	24/11/2022
		F19	Surgical Intervention	A0709	Device Sensing Problem	24/11/2022	24/11/2022
				A0712	Pacing Problem	24/11/2022	24/11/2022
CT2750	Implantable cardiac electrical stimulators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A0709	Device Sensing Problem	24/11/2022	24/11/2022
				A0712	Pacing Problem	24/11/2022	24/11/2022
		F08	Hospitalization or Prolonged Hospitalization	A0709	Device Sensing Problem	24/11/2022	24/11/2022
				A0712	Pacing Problem	24/11/2022	24/11/2022
		F19	Surgical Intervention	A0709	Device Sensing Problem	24/11/2022	24/11/2022
				A0712	Pacing Problem	24/11/2022	24/11/2022
CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A0709	Device Sensing Problem	24/11/2022	24/11/2022
				A0712	Pacing Problem	24/11/2022	24/11/2022
		F08	Hospitalization or Prolonged Hospitalization	A0709	Device Sensing Problem	24/11/2022	24/11/2022
				A0712	Pacing Problem	24/11/2022	24/11/2022
		F19	Surgical Intervention	A0709	Device Sensing Problem	24/11/2022	24/11/2022
				A0712	Pacing Problem	24/11/2022	24/11/2022
CT2985	Active	E2403	No Clinical Signs, Symptoms or Conditions	A0709	Device Sensing Problem	24/11/2022	24/11/2022
				A0712	Pacing Problem	24/11/2022	24/11/2022
		F08	Hospitalization or Prolonged Hospitalization	A0709	Device Sensing Problem	24/11/2022	24/11/2022
				A0712	Pacing Problem	24/11/2022	24/11/2022
		F19	Surgical Intervention	A0709	Device Sensing Problem	24/11/2022	24/11/2022
				A0712	Pacing Problem	24/11/2022	24/11/2022
CT3001	Implantable metal	E2403	No Clinical Signs, Symptoms or Conditions	A0709	Device Sensing Problem	24/11/2022	24/11/2022
				A0712	Pacing Problem	24/11/2022	24/11/2022
		F08	Hospitalization or Prolonged Hospitalization	A0709	Device Sensing Problem	24/11/2022	24/11/2022
				A0712	Pacing Problem	24/11/2022	24/11/2022
		F19	Surgical Intervention	A0709	Device Sensing Problem	24/11/2022	24/11/2022
				A0712	Pacing Problem	24/11/2022	24/11/2022
CT3002	Implantable synthetic polymer	E2403	No Clinical Signs, Symptoms or Conditions	A0709	Device Sensing Problem	24/11/2022	24/11/2022
				A0712	Pacing Problem	24/11/2022	24/11/2022
		F08	Hospitalization or Prolonged Hospitalization	A0709	Device Sensing Problem	24/11/2022	24/11/2022
				A0712	Pacing Problem	24/11/2022	24/11/2022
		F19	Surgical Intervention	A0709	Device Sensing Problem	24/11/2022	24/11/2022
				A0712	Pacing Problem	24/11/2022	24/11/2022
CT3006	Implantable inorganic material	E2403	No Clinical Signs, Symptoms or Conditions	A0709	Device Sensing Problem	24/11/2022	24/11/2022
				A0712	Pacing Problem	24/11/2022	24/11/2022
		F08	Hospitalization or Prolonged Hospitalization	A0709	Device Sensing Problem	24/11/2022	24/11/2022
				A0712	Pacing Problem	24/11/2022	24/11/2022
		F19	Surgical Intervention	A0709	Device Sensing Problem	24/11/2022	24/11/2022
				A0712	Pacing Problem	24/11/2022	24/11/2022
CT314	Electrically powered	E2403	No Clinical Signs, Symptoms or Conditions	A0709	Device Sensing Problem	24/11/2022	24/11/2022
				A0712	Pacing Problem	24/11/2022	24/11/2022
		F08	Hospitalization or Prolonged Hospitalization	A0709	Device Sensing Problem	24/11/2022	24/11/2022
				A0712	Pacing Problem	24/11/2022	24/11/2022
		F19	Surgical Intervention	A0709	Device Sensing Problem	24/11/2022	24/11/2022
				A0712	Pacing Problem	24/11/2022	24/11/2022
CT321	Long-term surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A0709	Device Sensing Problem	24/11/2022	24/11/2022
				A0712	Pacing Problem	24/11/2022	24/11/2022
		F08	Hospitalization or Prolonged Hospitalization	A0709	Device Sensing Problem	24/11/2022	24/11/2022
				A0712	Pacing Problem	24/11/2022	24/11/2022
		F19	Surgical Intervention	A0709	Device Sensing Problem	24/11/2022	24/11/2022
				A0712	Pacing Problem	24/11/2022	24/11/2022
CT449	Generators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A0709	Device Sensing Problem	24/11/2022	24/11/2022
				A0712	Pacing Problem	24/11/2022	24/11/2022
		F08	Hospitalization or Prolonged Hospitalization	A0709	Device Sensing Problem	24/11/2022	24/11/2022
				A0712	Pacing Problem	24/11/2022	24/11/2022
		F19	Surgical Intervention	A0709	Device Sensing Problem	24/11/2022	24/11/2022
				A0712	Pacing Problem	24/11/2022	24/11/2022
CT981	Single use	E2403	No Clinical Signs, Symptoms or Conditions	A0709	Device Sensing Problem	24/11/2022	24/11/2022
				A0712	Pacing Problem	24/11/2022	24/11/2022
		F08	Hospitalization or Prolonged Hospitalization	A0709	Device Sensing Problem	24/11/2022	24/11/2022
				A0712	Pacing Problem	24/11/2022	24/11/2022
		F19	Surgical Intervention	A0709	Device Sensing Problem	24/11/2022	24/11/2022
				A0712	Pacing Problem	24/11/2022	24/11/2022
CT983	Surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A0709	Device Sensing Problem	24/11/2022	24/11/2022
				A0712	Pacing Problem	24/11/2022	24/11/2022

							F08	Hospitalization or Prolonged Hospitalization	A0709	Device Sensing Problem	24/11/2022	24/11/2022
									A0712	Pacing Problem	24/11/2022	24/11/2022
							F19	Surgical Intervention	A0709	Device Sensing Problem	24/11/2022	24/11/2022
									A0712	Pacing Problem	24/11/2022	24/11/2022
11665	2022	All other reportable incidents	47265	Dual-chamber implantable pacemaker, rate-responsive	CT141	Electrodes/Leads and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	16/11/2022	16/11/2022
									A0908	Incorrect, Inadequate or Imprecise Result or Readings	16/11/2022	16/11/2022
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	16/11/2022	16/11/2022
									A0908	Incorrect, Inadequate or Imprecise Result or Readings	16/11/2022	16/11/2022
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	16/11/2022	16/11/2022
									A0908	Incorrect, Inadequate or Imprecise Result or Readings	16/11/2022	16/11/2022
					CT145	Cardiovascular devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	16/11/2022	16/11/2022
									A0908	Incorrect, Inadequate or Imprecise Result or Readings	16/11/2022	16/11/2022
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	16/11/2022	16/11/2022
									A0908	Incorrect, Inadequate or Imprecise Result or Readings	16/11/2022	16/11/2022
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	16/11/2022	16/11/2022
									A0908	Incorrect, Inadequate or Imprecise Result or Readings	16/11/2022	16/11/2022
					CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	16/11/2022	16/11/2022
									A0908	Incorrect, Inadequate or Imprecise Result or Readings	16/11/2022	16/11/2022
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	16/11/2022	16/11/2022
									A0908	Incorrect, Inadequate or Imprecise Result or Readings	16/11/2022	16/11/2022
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	16/11/2022	16/11/2022
									A0908	Incorrect, Inadequate or Imprecise Result or Readings	16/11/2022	16/11/2022
					CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	16/11/2022	16/11/2022
									A0908	Incorrect, Inadequate or Imprecise Result or Readings	16/11/2022	16/11/2022
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	16/11/2022	16/11/2022
									A0908	Incorrect, Inadequate or Imprecise Result or Readings	16/11/2022	16/11/2022
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	16/11/2022	16/11/2022
									A0908	Incorrect, Inadequate or Imprecise Result or Readings	16/11/2022	16/11/2022
					CT233	Surgical	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	16/11/2022	16/11/2022
									A0908	Incorrect, Inadequate or Imprecise Result or Readings	16/11/2022	16/11/2022
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	16/11/2022	16/11/2022
									A0908	Incorrect, Inadequate or Imprecise Result or Readings	16/11/2022	16/11/2022
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	16/11/2022	16/11/2022
									A0908	Incorrect, Inadequate or Imprecise Result or Readings	16/11/2022	16/11/2022
					CT2406	Implantable	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	16/11/2022	16/11/2022
									A0908	Incorrect, Inadequate or Imprecise Result or Readings	16/11/2022	16/11/2022
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	16/11/2022	16/11/2022
									A0908	Incorrect, Inadequate or Imprecise Result or Readings	16/11/2022	16/11/2022
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	16/11/2022	16/11/2022
									A0908	Incorrect, Inadequate or Imprecise Result or Readings	16/11/2022	16/11/2022
					CT2407	Active implantable	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	16/11/2022	16/11/2022
									A0908	Incorrect, Inadequate or Imprecise Result or Readings	16/11/2022	16/11/2022
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	16/11/2022	16/11/2022
									A0908	Incorrect, Inadequate or Imprecise Result or Readings	16/11/2022	16/11/2022
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	16/11/2022	16/11/2022
									A0908	Incorrect, Inadequate or Imprecise Result or Readings	16/11/2022	16/11/2022
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	16/11/2022	16/11/2022
									A0908	Incorrect, Inadequate or Imprecise Result or Readings	16/11/2022	16/11/2022
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	16/11/2022	16/11/2022
									A0908	Incorrect, Inadequate or Imprecise Result or Readings	16/11/2022	16/11/2022
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	16/11/2022	16/11/2022
									A0908	Incorrect, Inadequate or Imprecise Result or Readings	16/11/2022	16/11/2022
					CT2727	Active implant devices/ancillaries	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	16/11/2022	16/11/2022
									A0908	Incorrect, Inadequate or Imprecise Result or Readings	16/11/2022	16/11/2022
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	16/11/2022	16/11/2022
									A0908	Incorrect, Inadequate or Imprecise Result or Readings	16/11/2022	16/11/2022
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	16/11/2022	16/11/2022
									A0908	Incorrect, Inadequate or Imprecise Result or Readings	16/11/2022	16/11/2022
					CT2750	Implantable cardiac electrical stimulators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	16/11/2022	16/11/2022
									A0908	Incorrect, Inadequate or Imprecise Result or Readings	16/11/2022	16/11/2022
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	16/11/2022	16/11/2022
									A0908	Incorrect, Inadequate or Imprecise Result or Readings	16/11/2022	16/11/2022
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	16/11/2022	16/11/2022
									A0908	Incorrect, Inadequate or Imprecise Result or Readings	16/11/2022	16/11/2022
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	16/11/2022	16/11/2022

									A0908	Incorrect, Inadequate or Imprecise Result or Readings	16/11/2022	16/11/2022
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	16/11/2022	16/11/2022
									A0908	Incorrect, Inadequate or Imprecise Result or Readings	16/11/2022	16/11/2022
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	16/11/2022	16/11/2022
									A0908	Incorrect, Inadequate or Imprecise Result or Readings	16/11/2022	16/11/2022
					CT2985	Active	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	16/11/2022	16/11/2022
									A0908	Incorrect, Inadequate or Imprecise Result or Readings	16/11/2022	16/11/2022
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	16/11/2022	16/11/2022
									A0908	Incorrect, Inadequate or Imprecise Result or Readings	16/11/2022	16/11/2022
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	16/11/2022	16/11/2022
									A0908	Incorrect, Inadequate or Imprecise Result or Readings	16/11/2022	16/11/2022
					CT3001	Implantable metal	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	16/11/2022	16/11/2022
									A0908	Incorrect, Inadequate or Imprecise Result or Readings	16/11/2022	16/11/2022
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	16/11/2022	16/11/2022
									A0908	Incorrect, Inadequate or Imprecise Result or Readings	16/11/2022	16/11/2022
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	16/11/2022	16/11/2022
									A0908	Incorrect, Inadequate or Imprecise Result or Readings	16/11/2022	16/11/2022
					CT3002	Implantable synthetic polymer	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	16/11/2022	16/11/2022
									A0908	Incorrect, Inadequate or Imprecise Result or Readings	16/11/2022	16/11/2022
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	16/11/2022	16/11/2022
									A0908	Incorrect, Inadequate or Imprecise Result or Readings	16/11/2022	16/11/2022
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	16/11/2022	16/11/2022
									A0908	Incorrect, Inadequate or Imprecise Result or Readings	16/11/2022	16/11/2022
					CT3006	Implantable inorganic material	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	16/11/2022	16/11/2022
									A0908	Incorrect, Inadequate or Imprecise Result or Readings	16/11/2022	16/11/2022
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	16/11/2022	16/11/2022
									A0908	Incorrect, Inadequate or Imprecise Result or Readings	16/11/2022	16/11/2022
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	16/11/2022	16/11/2022
									A0908	Incorrect, Inadequate or Imprecise Result or Readings	16/11/2022	16/11/2022
					CT314	Electrically powered	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	16/11/2022	16/11/2022
									A0908	Incorrect, Inadequate or Imprecise Result or Readings	16/11/2022	16/11/2022
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	16/11/2022	16/11/2022
									A0908	Incorrect, Inadequate or Imprecise Result or Readings	16/11/2022	16/11/2022
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	16/11/2022	16/11/2022
									A0908	Incorrect, Inadequate or Imprecise Result or Readings	16/11/2022	16/11/2022
					CT321	Long-term surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	16/11/2022	16/11/2022
									A0908	Incorrect, Inadequate or Imprecise Result or Readings	16/11/2022	16/11/2022
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	16/11/2022	16/11/2022
									A0908	Incorrect, Inadequate or Imprecise Result or Readings	16/11/2022	16/11/2022
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	16/11/2022	16/11/2022
									A0908	Incorrect, Inadequate or Imprecise Result or Readings	16/11/2022	16/11/2022
					CT449	Generators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	16/11/2022	16/11/2022
									A0908	Incorrect, Inadequate or Imprecise Result or Readings	16/11/2022	16/11/2022
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	16/11/2022	16/11/2022
									A0908	Incorrect, Inadequate or Imprecise Result or Readings	16/11/2022	16/11/2022
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	16/11/2022	16/11/2022
									A0908	Incorrect, Inadequate or Imprecise Result or Readings	16/11/2022	16/11/2022
					CT981	Single use	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	16/11/2022	16/11/2022
									A0908	Incorrect, Inadequate or Imprecise Result or Readings	16/11/2022	16/11/2022
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	16/11/2022	16/11/2022
									A0908	Incorrect, Inadequate or Imprecise Result or Readings	16/11/2022	16/11/2022
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	16/11/2022	16/11/2022
									A0908	Incorrect, Inadequate or Imprecise Result or Readings	16/11/2022	16/11/2022
					CT983	Surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	16/11/2022	16/11/2022
									A0908	Incorrect, Inadequate or Imprecise Result or Readings	16/11/2022	16/11/2022
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	16/11/2022	16/11/2022
									A0908	Incorrect, Inadequate or Imprecise Result or Readings	16/11/2022	16/11/2022
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	16/11/2022	16/11/2022
									A0908	Incorrect, Inadequate or Imprecise Result or Readings	16/11/2022	16/11/2022
11666	2022	All other reportable incidents	17882	Physiologic-monitoring defibrillation system	CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070908	Failure to Sense		
							F26	No Health Consequences or Impact	A070908	Failure to Sense		
11667	2022	All other reportable incidents	36241	Pacemaker/defibrillator lead	CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A010402	Migration		
									A070909	Oversensing		
							F26	No Health Consequences or Impact	A010402	Migration		
									A070909	Oversensing		

11668	2022	All other reportable incidents	47263	Cardiac resynchronization therapy implantable pacemaker	CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A0908	Incorrect, Inadequate or Imprecise Result or Readings	09/12/2022	09/12/2022
							F08	Hospitalization or Prolonged Hospitalization	A0908	Incorrect, Inadequate or Imprecise Result or Readings	09/12/2022	09/12/2022
							F1905	Device Revision or Replacement	A0908	Incorrect, Inadequate or Imprecise Result or Readings	09/12/2022	09/12/2022
					CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A0908	Incorrect, Inadequate or Imprecise Result or Readings	09/12/2022	09/12/2022
							F08	Hospitalization or Prolonged Hospitalization	A0908	Incorrect, Inadequate or Imprecise Result or Readings	09/12/2022	09/12/2022
							F1905	Device Revision or Replacement	A0908	Incorrect, Inadequate or Imprecise Result or Readings	09/12/2022	09/12/2022
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A0908	Incorrect, Inadequate or Imprecise Result or Readings	09/12/2022	09/12/2022
							F08	Hospitalization or Prolonged Hospitalization	A0908	Incorrect, Inadequate or Imprecise Result or Readings	09/12/2022	09/12/2022
							F1905	Device Revision or Replacement	A0908	Incorrect, Inadequate or Imprecise Result or Readings	09/12/2022	09/12/2022
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A0908	Incorrect, Inadequate or Imprecise Result or Readings	09/12/2022	09/12/2022
							F08	Hospitalization or Prolonged Hospitalization	A0908	Incorrect, Inadequate or Imprecise Result or Readings	09/12/2022	09/12/2022
							F1905	Device Revision or Replacement	A0908	Incorrect, Inadequate or Imprecise Result or Readings	09/12/2022	09/12/2022
11669	2022		47265	Dual-chamber implantable pacemaker, rate-responsive	CT141	Electrodes/Leads and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A25	No Apparent Adverse Event	18/11/2022	18/11/2022
							F1901	Additional Surgery	A25	No Apparent Adverse Event	18/11/2022	18/11/2022
					CT145	Cardiovascular devices	E2403	No Clinical Signs, Symptoms or Conditions	A25	No Apparent Adverse Event	18/11/2022	18/11/2022
							F1901	Additional Surgery	A25	No Apparent Adverse Event	18/11/2022	18/11/2022
					CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A25	No Apparent Adverse Event	18/11/2022	18/11/2022
							F1901	Additional Surgery	A25	No Apparent Adverse Event	18/11/2022	18/11/2022
					CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A25	No Apparent Adverse Event	18/11/2022	18/11/2022
							F1901	Additional Surgery	A25	No Apparent Adverse Event	18/11/2022	18/11/2022
					CT233	Surgical	E2403	No Clinical Signs, Symptoms or Conditions	A25	No Apparent Adverse Event	18/11/2022	18/11/2022
							F1901	Additional Surgery	A25	No Apparent Adverse Event	18/11/2022	18/11/2022
					CT2406	Implantable	E2403	No Clinical Signs, Symptoms or Conditions	A25	No Apparent Adverse Event	18/11/2022	18/11/2022
							F1901	Additional Surgery	A25	No Apparent Adverse Event	18/11/2022	18/11/2022
					CT2407	Active implantable	E2403	No Clinical Signs, Symptoms or Conditions	A25	No Apparent Adverse Event	18/11/2022	18/11/2022
							F1901	Additional Surgery	A25	No Apparent Adverse Event	18/11/2022	18/11/2022
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A25	No Apparent Adverse Event	18/11/2022	18/11/2022
							F1901	Additional Surgery	A25	No Apparent Adverse Event	18/11/2022	18/11/2022
					CT2727	Active implant devices/ancillaries	E2403	No Clinical Signs, Symptoms or Conditions	A25	No Apparent Adverse Event	18/11/2022	18/11/2022
							F1901	Additional Surgery	A25	No Apparent Adverse Event	18/11/2022	18/11/2022
					CT2750	Implantable cardiac electrical stimulators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A25	No Apparent Adverse Event	18/11/2022	18/11/2022
							F1901	Additional Surgery	A25	No Apparent Adverse Event	18/11/2022	18/11/2022
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A25	No Apparent Adverse Event	18/11/2022	18/11/2022
							F1901	Additional Surgery	A25	No Apparent Adverse Event	18/11/2022	18/11/2022
					CT2985	Active	E2403	No Clinical Signs, Symptoms or Conditions	A25	No Apparent Adverse Event	18/11/2022	18/11/2022
							F1901	Additional Surgery	A25	No Apparent Adverse Event	18/11/2022	18/11/2022
					CT3001	Implantable metal	E2403	No Clinical Signs, Symptoms or Conditions	A25	No Apparent Adverse Event	18/11/2022	18/11/2022
							F1901	Additional Surgery	A25	No Apparent Adverse Event	18/11/2022	18/11/2022
					CT3002	Implantable synthetic polymer	E2403	No Clinical Signs, Symptoms or Conditions	A25	No Apparent Adverse Event	18/11/2022	18/11/2022
							F1901	Additional Surgery	A25	No Apparent Adverse Event	18/11/2022	18/11/2022
					CT3006	Implantable inorganic material	E2403	No Clinical Signs, Symptoms or Conditions	A25	No Apparent Adverse Event	18/11/2022	18/11/2022
							F1901	Additional Surgery	A25	No Apparent Adverse Event	18/11/2022	18/11/2022
					CT314	Electrically powered	E2403	No Clinical Signs, Symptoms or Conditions	A25	No Apparent Adverse Event	18/11/2022	18/11/2022
							F1901	Additional Surgery	A25	No Apparent Adverse Event	18/11/2022	18/11/2022
					CT321	Long-term surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A25	No Apparent Adverse Event	18/11/2022	18/11/2022
							F1901	Additional Surgery	A25	No Apparent Adverse Event	18/11/2022	18/11/2022
					CT449	Generators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A25	No Apparent Adverse Event	18/11/2022	18/11/2022
							F1901	Additional Surgery	A25	No Apparent Adverse Event	18/11/2022	18/11/2022
					CT981	Single use	E2403	No Clinical Signs, Symptoms or Conditions	A25	No Apparent Adverse Event	18/11/2022	18/11/2022
							F1901	Additional Surgery	A25	No Apparent Adverse Event	18/11/2022	18/11/2022
					CT983	Surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A25	No Apparent Adverse Event	18/11/2022	18/11/2022
							F1901	Additional Surgery	A25	No Apparent Adverse Event	18/11/2022	18/11/2022
11670	2022		35223	Endocardial/interventricular septal pacing lead	CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070909	Oversensing	12/10/2022	12/10/2022
							F1908	Prolonged Surgery	A070909	Oversensing	12/10/2022	12/10/2022
11671	2022	Unanticipated serious deterioration in state of health	35223	Endocardial/interventricular septal pacing lead	CT1684	Cardiac pacemakers and associated devices			A0101	Patient-Device Incompatibility		
11672	2022	All other reportable incidents	47265	Dual-chamber implantable pacemaker, rate-responsive	CT141	Electrodes/Leads and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070101	Failure to Capture		
									A070910	Undersensing		
									A0909	Unexpected Therapeutic Results		
							F26	No Health Consequences or Impact	A070101	Failure to Capture		
									A070910	Undersensing		
									A0909	Unexpected Therapeutic Results		
					CT145	Cardiovascular devices	E2403	No Clinical Signs, Symptoms or Conditions	A070101	Failure to Capture		
									A070910	Undersensing		
									A0909	Unexpected Therapeutic Results		
							F26	No Health Consequences or Impact	A070101	Failure to Capture		
									A070910	Undersensing		
									A0909	Unexpected Therapeutic Results		
					CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070101	Failure to Capture		
									A070910	Undersensing		
									A0909	Unexpected Therapeutic Results		
							F26	No Health Consequences or Impact	A070101	Failure to Capture		
									A070910	Undersensing		
									A0909	Unexpected Therapeutic Results		
					CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A070101	Failure to Capture		
									A070910	Undersensing		
									A0909	Unexpected Therapeutic Results		
							F26	No Health Consequences or Impact	A070101	Failure to Capture		
									A070910	Undersensing		

								A0909	Unexpected Therapeutic Results		
					CT233	Surgical	E2403	No Clinical Signs, Symptoms or Conditions	A070101	Failure to Capture		
									A070910	Undersensing		
									A0909	Unexpected Therapeutic Results		
							F26	No Health Consequences or Impact	A070101	Failure to Capture		
									A070910	Undersensing		
									A0909	Unexpected Therapeutic Results		
					CT2406	Implantable	E2403	No Clinical Signs, Symptoms or Conditions	A070101	Failure to Capture		
									A070910	Undersensing		
									A0909	Unexpected Therapeutic Results		
							F26	No Health Consequences or Impact	A070101	Failure to Capture		
									A070910	Undersensing		
									A0909	Unexpected Therapeutic Results		
					CT2407	Active implantable	E2403	No Clinical Signs, Symptoms or Conditions	A070101	Failure to Capture		
									A070910	Undersensing		
									A0909	Unexpected Therapeutic Results		
							F26	No Health Consequences or Impact	A070101	Failure to Capture		
									A070910	Undersensing		
									A0909	Unexpected Therapeutic Results		
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A070101	Failure to Capture		
									A070910	Undersensing		
									A0909	Unexpected Therapeutic Results		
							F26	No Health Consequences or Impact	A070101	Failure to Capture		
									A070910	Undersensing		
									A0909	Unexpected Therapeutic Results		
					CT2727	Active implant devices/ancillaries	E2403	No Clinical Signs, Symptoms or Conditions	A070101	Failure to Capture		
									A070910	Undersensing		
									A0909	Unexpected Therapeutic Results		
							F26	No Health Consequences or Impact	A070101	Failure to Capture		
									A070910	Undersensing		
									A0909	Unexpected Therapeutic Results		
					CT2750	Implantable cardiac electrical stimulators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070101	Failure to Capture		
									A070910	Undersensing		
									A0909	Unexpected Therapeutic Results		
							F26	No Health Consequences or Impact	A070101	Failure to Capture		
									A070910	Undersensing		
									A0909	Unexpected Therapeutic Results		
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A070101	Failure to Capture		
									A070910	Undersensing		
									A0909	Unexpected Therapeutic Results		
							F26	No Health Consequences or Impact	A070101	Failure to Capture		
									A070910	Undersensing		
									A0909	Unexpected Therapeutic Results		
					CT2985	Active	E2403	No Clinical Signs, Symptoms or Conditions	A070101	Failure to Capture		
									A070910	Undersensing		
									A0909	Unexpected Therapeutic Results		
							F26	No Health Consequences or Impact	A070101	Failure to Capture		
									A070910	Undersensing		
									A0909	Unexpected Therapeutic Results		
					CT3001	Implantable metal	E2403	No Clinical Signs, Symptoms or Conditions	A070101	Failure to Capture		
									A070910	Undersensing		
									A0909	Unexpected Therapeutic Results		
							F26	No Health Consequences or Impact	A070101	Failure to Capture		
									A070910	Undersensing		
									A0909	Unexpected Therapeutic Results		
					CT3002	Implantable synthetic polymer	E2403	No Clinical Signs, Symptoms or Conditions	A070101	Failure to Capture		
									A070910	Undersensing		
									A0909	Unexpected Therapeutic Results		
							F26	No Health Consequences or Impact	A070101	Failure to Capture		
									A070910	Undersensing		
									A0909	Unexpected Therapeutic Results		
					CT3006	Implantable inorganic material	E2403	No Clinical Signs, Symptoms or Conditions	A070101	Failure to Capture		
									A070910	Undersensing		
									A0909	Unexpected Therapeutic Results		
							F26	No Health Consequences or Impact	A070101	Failure to Capture		
									A070910	Undersensing		
									A0909	Unexpected Therapeutic Results		
					CT314	Electrically powered	E2403	No Clinical Signs, Symptoms or Conditions	A070101	Failure to Capture		
									A070910	Undersensing		
									A0909	Unexpected Therapeutic Results		
							F26	No Health Consequences or Impact	A070101	Failure to Capture		
									A070910	Undersensing		
									A0909	Unexpected Therapeutic Results		
					CT321	Long-term surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A070101	Failure to Capture		
									A070910	Undersensing		
									A0909	Unexpected Therapeutic Results		
							F26	No Health Consequences or Impact	A070101	Failure to Capture		
									A070910	Undersensing		
									A0909	Unexpected Therapeutic Results		
					CT449	Generators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070101	Failure to Capture		
									A070910	Undersensing		
									A0909	Unexpected Therapeutic Results		
							F26	No Health Consequences or Impact	A070101	Failure to Capture		
									A070910	Undersensing		
									A0909	Unexpected Therapeutic Results		
					CT981	Single use	E2403	No Clinical Signs, Symptoms or Conditions	A070101	Failure to Capture		
									A070910	Undersensing		
									A0909	Unexpected Therapeutic Results		

							F26	No Health Consequences or Impact	A070101	Failure to Capture		
									A070910	Undersensing		
									A0909	Unexpected Therapeutic Results		
					CT983	Surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A070101	Failure to Capture		
									A070910	Undersensing		
									A0909	Unexpected Therapeutic Results		
							F26	No Health Consequences or Impact	A070101	Failure to Capture		
									A070910	Undersensing		
									A0909	Unexpected Therapeutic Results		
11673	2022	All other reportable incidents	35223	Endocardial/interventricular septal pacing lead	CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070909	Oversensing		
									A0722	Impedance Problem		
									A090801	Signal Artifact/Noise		
							F26	No Health Consequences or Impact	A070909	Oversensing		
									A0722	Impedance Problem		
									A090801	Signal Artifact/Noise		
11674	2022	All other reportable incidents	35223	Endocardial/interventricular septal pacing lead	CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070104	Unstable Capture Threshold		
									A070909	Oversensing		
							F26	No Health Consequences or Impact	A070104	Unstable Capture Threshold		
									A070909	Oversensing		
11675	2022	All other reportable incidents	17882	Physiologic-monitoring defibrillation system	CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070803	Failure to Power Up		
							F27	Problem Identified During Non-Clinical Procedure	A070803	Failure to Power Up		
11676	2022	All other reportable incidents	35852	Single-chamber implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
									A0908	Incorrect, Inadequate or Imprecise Result or Readings		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
									A0908	Incorrect, Inadequate or Imprecise Result or Readings		
11677	2022	All other reportable incidents	36241	Pacemaker/defibrillator lead	CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070909	Oversensing		
									A0712	Pacing Problem		
							F10	Inadequate/inappropriate Treatment or Diagnostic Exposure	A070909	Oversensing		
									A0712	Pacing Problem		
							F2203	Imaging Required	A070909	Oversensing		
									A0712	Pacing Problem		
11678	2022		47265	Dual-chamber implantable pacemaker, rate-responsive	CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	23/11/2022	23/11/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	23/11/2022	23/11/2022
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	23/11/2022	23/11/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	23/11/2022	23/11/2022
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	23/11/2022	23/11/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	23/11/2022	23/11/2022
11679	2022		47265	Dual-chamber implantable pacemaker, rate-responsive	CT141	Electrodes/Leads and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	06/10/2022	06/10/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	06/10/2022	06/10/2022
					CT145	Cardiovascular devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	06/10/2022	06/10/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	06/10/2022	06/10/2022
					CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	06/10/2022	06/10/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	06/10/2022	06/10/2022
					CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	06/10/2022	06/10/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	06/10/2022	06/10/2022
					CT233	Surgical	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	06/10/2022	06/10/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	06/10/2022	06/10/2022
					CT2406	Implantable	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	06/10/2022	06/10/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	06/10/2022	06/10/2022
					CT2407	Active implantable	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	06/10/2022	06/10/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	06/10/2022	06/10/2022
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	06/10/2022	06/10/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	06/10/2022	06/10/2022
					CT2727	Active implant devices/ancillaries	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	06/10/2022	06/10/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	06/10/2022	06/10/2022
					CT2750	Implantable cardiac electrical stimulators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	06/10/2022	06/10/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	06/10/2022	06/10/2022
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	06/10/2022	06/10/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	06/10/2022	06/10/2022
					CT2985	Active	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	06/10/2022	06/10/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	06/10/2022	06/10/2022
					CT3001	Implantable metal	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	06/10/2022	06/10/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	06/10/2022	06/10/2022
					CT3002	Implantable synthetic polymer	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	06/10/2022	06/10/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	06/10/2022	06/10/2022
					CT3006	Implantable inorganic material	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	06/10/2022	06/10/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	06/10/2022	06/10/2022
					CT314	Electrically powered	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	06/10/2022	06/10/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	06/10/2022	06/10/2022
					CT321	Long-term surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	06/10/2022	06/10/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	06/10/2022	06/10/2022
					CT449	Generators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	06/10/2022	06/10/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	06/10/2022	06/10/2022
					CT981	Single use	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	06/10/2022	06/10/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	06/10/2022	06/10/2022
					CT983	Surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	06/10/2022	06/10/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	06/10/2022	06/10/2022
11680	2022		47265	Dual-chamber implantable pacemaker, rate-responsive	CT141	Electrodes/Leads and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	11/11/2022	11/11/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	11/11/2022	11/11/2022
					CT145	Cardiovascular devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	11/11/2022	11/11/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	11/11/2022	11/11/2022
					CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	11/11/2022	11/11/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	11/11/2022	11/11/2022
					CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	11/11/2022	11/11/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	11/11/2022	11/11/2022
					CT233	Surgical	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	11/11/2022	11/11/2022

							F1901	Additional Surgery	A24	No Apparent Device Problem	11/11/2022	11/11/2022
					CT2406	Implantable	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	11/11/2022	11/11/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	11/11/2022	11/11/2022
					CT2407	Active implantable	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	11/11/2022	11/11/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	11/11/2022	11/11/2022
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	11/11/2022	11/11/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	11/11/2022	11/11/2022
					CT2727	Active implant devices/ancillaries	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	11/11/2022	11/11/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	11/11/2022	11/11/2022
					CT2750	Implantable cardiac electrical stimulators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	11/11/2022	11/11/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	11/11/2022	11/11/2022
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	11/11/2022	11/11/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	11/11/2022	11/11/2022
					CT2985	Active	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	11/11/2022	11/11/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	11/11/2022	11/11/2022
					CT3001	Implantable metal	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	11/11/2022	11/11/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	11/11/2022	11/11/2022
					CT3002	Implantable synthetic polymer	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	11/11/2022	11/11/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	11/11/2022	11/11/2022
					CT3006	Implantable inorganic material	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	11/11/2022	11/11/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	11/11/2022	11/11/2022
					CT314	Electrically powered	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	11/11/2022	11/11/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	11/11/2022	11/11/2022
					CT321	Long-term surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	11/11/2022	11/11/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	11/11/2022	11/11/2022
					CT449	Generators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	11/11/2022	11/11/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	11/11/2022	11/11/2022
					CT981	Single use	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	11/11/2022	11/11/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	11/11/2022	11/11/2022
					CT983	Surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	11/11/2022	11/11/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	11/11/2022	11/11/2022
11681	2022		47265	Dual-chamber implantable pacemaker, rate-responsive	CT141	Electrodes/Leads and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	06/10/2022	06/10/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	06/10/2022	06/10/2022
					CT145	Cardiovascular devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	06/10/2022	06/10/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	06/10/2022	06/10/2022
					CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	06/10/2022	06/10/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	06/10/2022	06/10/2022
					CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	06/10/2022	06/10/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	06/10/2022	06/10/2022
					CT233	Surgical	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	06/10/2022	06/10/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	06/10/2022	06/10/2022
					CT2406	Implantable	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	06/10/2022	06/10/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	06/10/2022	06/10/2022
					CT2407	Active implantable	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	06/10/2022	06/10/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	06/10/2022	06/10/2022
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	06/10/2022	06/10/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	06/10/2022	06/10/2022
					CT2727	Active implant devices/ancillaries	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	06/10/2022	06/10/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	06/10/2022	06/10/2022
					CT2750	Implantable cardiac electrical stimulators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	06/10/2022	06/10/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	06/10/2022	06/10/2022
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	06/10/2022	06/10/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	06/10/2022	06/10/2022
					CT2985	Active	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	06/10/2022	06/10/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	06/10/2022	06/10/2022
					CT3001	Implantable metal	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	06/10/2022	06/10/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	06/10/2022	06/10/2022
					CT3002	Implantable synthetic polymer	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	06/10/2022	06/10/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	06/10/2022	06/10/2022
					CT3006	Implantable inorganic material	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	06/10/2022	06/10/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	06/10/2022	06/10/2022
					CT314	Electrically powered	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	06/10/2022	06/10/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	06/10/2022	06/10/2022
					CT321	Long-term surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	06/10/2022	06/10/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	06/10/2022	06/10/2022
					CT449	Generators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	06/10/2022	06/10/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	06/10/2022	06/10/2022
					CT981	Single use	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	06/10/2022	06/10/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	06/10/2022	06/10/2022
					CT983	Surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	06/10/2022	06/10/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	06/10/2022	06/10/2022
11682	2022		47265	Dual-chamber implantable pacemaker, rate-responsive	CT141	Electrodes/Leads and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem		
							F1901	Additional Surgery	A24	No Apparent Device Problem		
					CT145	Cardiovascular devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem		
							F1901	Additional Surgery	A24	No Apparent Device Problem		
					CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem		
							F1901	Additional Surgery	A24	No Apparent Device Problem		
					CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem		
							F1901	Additional Surgery	A24	No Apparent Device Problem		
					CT233	Surgical	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem		
							F1901	Additional Surgery	A24	No Apparent Device Problem		
					CT2406	Implantable	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem		
							F1901	Additional Surgery	A24	No Apparent Device Problem		
					CT2407	Active implantable	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem		
							F1901	Additional Surgery	A24	No Apparent Device Problem		
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem		
							F1901	Additional Surgery	A24	No Apparent Device Problem		
					CT2727	Active implant devices/ancillaries	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem		

							F1901	Additional Surgery	A24	No Apparent Device Problem		
					CT2750	Implantable cardiac electrical stimulators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem		
							F1901	Additional Surgery	A24	No Apparent Device Problem		
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem		
							F1901	Additional Surgery	A24	No Apparent Device Problem		
					CT2985	Active	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem		
							F1901	Additional Surgery	A24	No Apparent Device Problem		
					CT3001	Implantable metal	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem		
							F1901	Additional Surgery	A24	No Apparent Device Problem		
					CT3002	Implantable synthetic polymer	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem		
							F1901	Additional Surgery	A24	No Apparent Device Problem		
					CT3006	Implantable inorganic material	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem		
							F1901	Additional Surgery	A24	No Apparent Device Problem		
					CT314	Electrically powered	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem		
							F1901	Additional Surgery	A24	No Apparent Device Problem		
					CT321	Long-term surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem		
							F1901	Additional Surgery	A24	No Apparent Device Problem		
					CT449	Generators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem		
							F1901	Additional Surgery	A24	No Apparent Device Problem		
					CT981	Single use	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem		
							F1901	Additional Surgery	A24	No Apparent Device Problem		
					CT983	Surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem		
							F1901	Additional Surgery	A24	No Apparent Device Problem		
11683	2022		47265	Dual-chamber implantable pacemaker, rate-responsive	CT141	Electrodes/Leads and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	21/11/2022	21/11/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	21/11/2022	21/11/2022
					CT145	Cardiovascular devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	21/11/2022	21/11/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	21/11/2022	21/11/2022
					CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	21/11/2022	21/11/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	21/11/2022	21/11/2022
					CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	21/11/2022	21/11/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	21/11/2022	21/11/2022
					CT233	Surgical	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	21/11/2022	21/11/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	21/11/2022	21/11/2022
					CT2406	Implantable	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	21/11/2022	21/11/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	21/11/2022	21/11/2022
					CT2407	Active implantable	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	21/11/2022	21/11/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	21/11/2022	21/11/2022
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	21/11/2022	21/11/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	21/11/2022	21/11/2022
					CT2727	Active implant devices/ancillaries	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	21/11/2022	21/11/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	21/11/2022	21/11/2022
					CT2750	Implantable cardiac electrical stimulators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	21/11/2022	21/11/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	21/11/2022	21/11/2022
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	21/11/2022	21/11/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	21/11/2022	21/11/2022
					CT2985	Active	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	21/11/2022	21/11/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	21/11/2022	21/11/2022
					CT3001	Implantable metal	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	21/11/2022	21/11/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	21/11/2022	21/11/2022
					CT3002	Implantable synthetic polymer	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	21/11/2022	21/11/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	21/11/2022	21/11/2022
					CT3006	Implantable inorganic material	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	21/11/2022	21/11/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	21/11/2022	21/11/2022
					CT314	Electrically powered	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	21/11/2022	21/11/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	21/11/2022	21/11/2022
					CT321	Long-term surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	21/11/2022	21/11/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	21/11/2022	21/11/2022
					CT449	Generators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	21/11/2022	21/11/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	21/11/2022	21/11/2022
					CT981	Single use	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	21/11/2022	21/11/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	21/11/2022	21/11/2022
					CT983	Surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	21/11/2022	21/11/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	21/11/2022	21/11/2022
11684	2022	All other reportable incidents	47263	Cardiac resynchronization therapy implantable pacemaker	CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070102	High Capture Threshold		
									A070909	Oversensing		
									A072201	High Impedance		
									A090801	Signal Artifact/Noise		
							F26	No Health Consequences or Impact	A070102	High Capture Threshold		
									A070909	Oversensing		
									A072201	High Impedance		
									A090801	Signal Artifact/Noise		
					CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A070102	High Capture Threshold		
									A070909	Oversensing		
									A072201	High Impedance		
									A090801	Signal Artifact/Noise		
							F26	No Health Consequences or Impact	A070102	High Capture Threshold		
									A070909	Oversensing		
									A072201	High Impedance		
									A090801	Signal Artifact/Noise		
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A070102	High Capture Threshold		
									A070909	Oversensing		
									A072201	High Impedance		
									A090801	Signal Artifact/Noise		
							F26	No Health Consequences or Impact	A070102	High Capture Threshold		
									A070909	Oversensing		
									A072201	High Impedance		
									A090801	Signal Artifact/Noise		
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A070102	High Capture Threshold		

									A070909	Oversensing		
									A072201	High Impedance		
									A090801	Signal Artifact/Noise		
							F26	No Health Consequences or Impact	A070102	High Capture Threshold		
									A070909	Oversensing		
									A072201	High Impedance		
									A090801	Signal Artifact/Noise		
11685	2022	All other reportable incidents	35223	Endocardial/interventricular septal pacing lead	CT1684	Cardiac pacemakers and associated devices	E2103	Device Overstimulation of Tissue	A24	No Apparent Device Problem		
							E2311	Discomfort	A24	No Apparent Device Problem		
							F12	Serious Injury/ Illness/ Impairment	A24	No Apparent Device Problem		
							F23	Unexpected Medical Intervention	A24	No Apparent Device Problem		
11686	2022	All other reportable incidents	47265	Dual-chamber implantable pacemaker, rate-responsive	CT141	Electrodes/Leads and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F24	Insufficient Health Impact Information	A070504	Premature Discharge of Battery		
					CT145	Cardiovascular devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F24	Insufficient Health Impact Information	A070504	Premature Discharge of Battery		
					CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F24	Insufficient Health Impact Information	A070504	Premature Discharge of Battery		
					CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F24	Insufficient Health Impact Information	A070504	Premature Discharge of Battery		
					CT233	Surgical	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F24	Insufficient Health Impact Information	A070504	Premature Discharge of Battery		
					CT2406	Implantable	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F24	Insufficient Health Impact Information	A070504	Premature Discharge of Battery		
					CT2407	Active implantable	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F24	Insufficient Health Impact Information	A070504	Premature Discharge of Battery		
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F24	Insufficient Health Impact Information	A070504	Premature Discharge of Battery		
					CT2727	Active implant devices/ancillaries	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F24	Insufficient Health Impact Information	A070504	Premature Discharge of Battery		
					CT2750	Implantable cardiac electrical stimulators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F24	Insufficient Health Impact Information	A070504	Premature Discharge of Battery		
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F24	Insufficient Health Impact Information	A070504	Premature Discharge of Battery		
					CT2985	Active	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F24	Insufficient Health Impact Information	A070504	Premature Discharge of Battery		
					CT3001	Implantable metal	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F24	Insufficient Health Impact Information	A070504	Premature Discharge of Battery		
					CT3002	Implantable synthetic polymer	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F24	Insufficient Health Impact Information	A070504	Premature Discharge of Battery		
					CT3006	Implantable inorganic material	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F24	Insufficient Health Impact Information	A070504	Premature Discharge of Battery		
					CT314	Electrically powered	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F24	Insufficient Health Impact Information	A070504	Premature Discharge of Battery		
					CT321	Long-term surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F24	Insufficient Health Impact Information	A070504	Premature Discharge of Battery		
					CT449	Generators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F24	Insufficient Health Impact Information	A070504	Premature Discharge of Battery		
					CT981	Single use	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F24	Insufficient Health Impact Information	A070504	Premature Discharge of Battery		
					CT983	Surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F24	Insufficient Health Impact Information	A070504	Premature Discharge of Battery		
11687	2022	All other reportable incidents	60910	Coronary venous pacing lead	CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A072201	High Impedance		
							F23	Unexpected Medical Intervention	A072201	High Impedance		
11688	2022	All other reportable incidents	47265	Dual-chamber implantable pacemaker, rate-responsive	CT141	Electrodes/Leads and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT145	Cardiovascular devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT233	Surgical	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT2406	Implantable	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT2407	Active implantable	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT2727	Active implant devices/ancillaries	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT2750	Implantable cardiac electrical stimulators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT2985	Active	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT3001	Implantable metal	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT3002	Implantable synthetic polymer	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT3006	Implantable inorganic material	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT314	Electrically powered	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT321	Long-term surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT449	Generators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		

						F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
				CT981	Single use	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
						F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
				CT983	Surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
						F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
11689	2022	All other reportable incidents	47265	Dual-chamber implantable pacemaker, rate-responsive	CT141	Electrodes/Leads and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	07/11/2022	07/11/2022
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	07/11/2022	07/11/2022
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	07/11/2022	07/11/2022
					CT145	Cardiovascular devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	07/11/2022	07/11/2022
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	07/11/2022	07/11/2022
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	07/11/2022	07/11/2022
					CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	07/11/2022	07/11/2022
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	07/11/2022	07/11/2022
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	07/11/2022	07/11/2022
					CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	07/11/2022	07/11/2022
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	07/11/2022	07/11/2022
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	07/11/2022	07/11/2022
					CT233	Surgical	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	07/11/2022	07/11/2022
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	07/11/2022	07/11/2022
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	07/11/2022	07/11/2022
					CT2406	Implantable	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	07/11/2022	07/11/2022
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	07/11/2022	07/11/2022
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	07/11/2022	07/11/2022
					CT2407	Active implantable	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	07/11/2022	07/11/2022
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	07/11/2022	07/11/2022
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	07/11/2022	07/11/2022
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	07/11/2022	07/11/2022
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	07/11/2022	07/11/2022
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	07/11/2022	07/11/2022
					CT2727	Active implant devices/ancillaries	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	07/11/2022	07/11/2022
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	07/11/2022	07/11/2022
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	07/11/2022	07/11/2022
					CT2750	Implantable cardiac electrical stimulators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	07/11/2022	07/11/2022
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	07/11/2022	07/11/2022
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	07/11/2022	07/11/2022
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	07/11/2022	07/11/2022
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	07/11/2022	07/11/2022
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	07/11/2022	07/11/2022
					CT2985	Active	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	07/11/2022	07/11/2022
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	07/11/2022	07/11/2022
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	07/11/2022	07/11/2022
					CT3001	Implantable metal	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	07/11/2022	07/11/2022
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	07/11/2022	07/11/2022
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	07/11/2022	07/11/2022
					CT3002	Implantable synthetic polymer	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	07/11/2022	07/11/2022
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	07/11/2022	07/11/2022
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	07/11/2022	07/11/2022
					CT3006	Implantable inorganic material	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	07/11/2022	07/11/2022
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	07/11/2022	07/11/2022
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	07/11/2022	07/11/2022
					CT314	Electrically powered	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	07/11/2022	07/11/2022
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	07/11/2022	07/11/2022
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	07/11/2022	07/11/2022
					CT321	Long-term surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	07/11/2022	07/11/2022
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	07/11/2022	07/11/2022
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	07/11/2022	07/11/2022
					CT449	Generators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	07/11/2022	07/11/2022
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	07/11/2022	07/11/2022
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	07/11/2022	07/11/2022
					CT981	Single use	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	07/11/2022	07/11/2022
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	07/11/2022	07/11/2022
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	07/11/2022	07/11/2022
					CT983	Surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	07/11/2022	07/11/2022
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	07/11/2022	07/11/2022
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	07/11/2022	07/11/2022
11690	2022	All other reportable incidents	35223	Endocardial/interventricular septal pacing lead	CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070101	Failure to Capture		
									A070909	Oversensing		
									A070910	Undersensing		
							F23	Unexpected Medical Intervention	A070101	Failure to Capture		
									A070909	Oversensing		
									A070910	Undersensing		
11691	2022	All other reportable incidents	37265	Dual-chamber implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A0712	Pacing Problem		
							F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A0712	Pacing Problem		
11692	2022	All other reportable incidents	17882	Physiologic-monitoring defibrillation system	CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A071302	Inappropriate/Inadequate Shock/Stimulation		
									A12	Connection Problem		
							F27	Problem Identified During Non-Clinical Procedure	A071302	Inappropriate/Inadequate Shock/Stimulation		
									A12	Connection Problem		
11693	2022	All other reportable incidents	47265	Dual-chamber implantable pacemaker, rate-responsive	CT141	Electrodes/Leads and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A071102	Failure to Interrogate	25/11/2022	25/11/2022
							F1901	Additional Surgery	A071102	Failure to Interrogate	25/11/2022	25/11/2022
					CT145	Cardiovascular devices	E2403	No Clinical Signs, Symptoms or Conditions	A071102	Failure to Interrogate	25/11/2022	25/11/2022
							F1901	Additional Surgery	A071102	Failure to Interrogate	25/11/2022	25/11/2022
					CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A071102	Failure to Interrogate	25/11/2022	25/11/2022
							F1901	Additional Surgery	A071102	Failure to Interrogate	25/11/2022	25/11/2022
					CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A071102	Failure to Interrogate	25/11/2022	25/11/2022
							F1901	Additional Surgery	A071102	Failure to Interrogate	25/11/2022	25/11/2022
					CT233	Surgical	E2403	No Clinical Signs, Symptoms or Conditions	A071102	Failure to Interrogate	25/11/2022	25/11/2022
							F1901	Additional Surgery	A071102	Failure to Interrogate	25/11/2022	25/11/2022
					CT2406	Implantable	E2403	No Clinical Signs, Symptoms or Conditions	A071102	Failure to Interrogate	25/11/2022	25/11/2022

							F1901	Additional Surgery	A071102	Failure to Interrogate	25/11/2022	25/11/2022
					CT2407	Active implantable	E2403	No Clinical Signs, Symptoms or Conditions	A071102	Failure to Interrogate	25/11/2022	25/11/2022
							F1901	Additional Surgery	A071102	Failure to Interrogate	25/11/2022	25/11/2022
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A071102	Failure to Interrogate	25/11/2022	25/11/2022
							F1901	Additional Surgery	A071102	Failure to Interrogate	25/11/2022	25/11/2022
					CT2727	Active implant devices/ancillaries	E2403	No Clinical Signs, Symptoms or Conditions	A071102	Failure to Interrogate	25/11/2022	25/11/2022
							F1901	Additional Surgery	A071102	Failure to Interrogate	25/11/2022	25/11/2022
					CT2750	Implantable cardiac electrical stimulators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A071102	Failure to Interrogate	25/11/2022	25/11/2022
							F1901	Additional Surgery	A071102	Failure to Interrogate	25/11/2022	25/11/2022
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A071102	Failure to Interrogate	25/11/2022	25/11/2022
							F1901	Additional Surgery	A071102	Failure to Interrogate	25/11/2022	25/11/2022
					CT2985	Active	E2403	No Clinical Signs, Symptoms or Conditions	A071102	Failure to Interrogate	25/11/2022	25/11/2022
							F1901	Additional Surgery	A071102	Failure to Interrogate	25/11/2022	25/11/2022
					CT3001	Implantable metal	E2403	No Clinical Signs, Symptoms or Conditions	A071102	Failure to Interrogate	25/11/2022	25/11/2022
							F1901	Additional Surgery	A071102	Failure to Interrogate	25/11/2022	25/11/2022
					CT3002	Implantable synthetic polymer	E2403	No Clinical Signs, Symptoms or Conditions	A071102	Failure to Interrogate	25/11/2022	25/11/2022
							F1901	Additional Surgery	A071102	Failure to Interrogate	25/11/2022	25/11/2022
					CT3006	Implantable inorganic material	E2403	No Clinical Signs, Symptoms or Conditions	A071102	Failure to Interrogate	25/11/2022	25/11/2022
							F1901	Additional Surgery	A071102	Failure to Interrogate	25/11/2022	25/11/2022
					CT314	Electrically powered	E2403	No Clinical Signs, Symptoms or Conditions	A071102	Failure to Interrogate	25/11/2022	25/11/2022
							F1901	Additional Surgery	A071102	Failure to Interrogate	25/11/2022	25/11/2022
					CT321	Long-term surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A071102	Failure to Interrogate	25/11/2022	25/11/2022
							F1901	Additional Surgery	A071102	Failure to Interrogate	25/11/2022	25/11/2022
					CT449	Generators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A071102	Failure to Interrogate	25/11/2022	25/11/2022
							F1901	Additional Surgery	A071102	Failure to Interrogate	25/11/2022	25/11/2022
					CT981	Single use	E2403	No Clinical Signs, Symptoms or Conditions	A071102	Failure to Interrogate	25/11/2022	25/11/2022
							F1901	Additional Surgery	A071102	Failure to Interrogate	25/11/2022	25/11/2022
					CT983	Surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A071102	Failure to Interrogate	25/11/2022	25/11/2022
							F1901	Additional Surgery	A071102	Failure to Interrogate	25/11/2022	25/11/2022
11694	2022	All other reportable incidents	47265	Dual-chamber implantable pacemaker, rate-responsive	CT141	Electrodes/Leads and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A071205	Pacemaker Found in Back-Up Mode		
							F26	No Health Consequences or Impact	A071205	Pacemaker Found in Back-Up Mode		
					CT145	Cardiovascular devices	E2403	No Clinical Signs, Symptoms or Conditions	A071205	Pacemaker Found in Back-Up Mode		
							F26	No Health Consequences or Impact	A071205	Pacemaker Found in Back-Up Mode		
					CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A071205	Pacemaker Found in Back-Up Mode		
							F26	No Health Consequences or Impact	A071205	Pacemaker Found in Back-Up Mode		
					CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A071205	Pacemaker Found in Back-Up Mode		
							F26	No Health Consequences or Impact	A071205	Pacemaker Found in Back-Up Mode		
					CT233	Surgical	E2403	No Clinical Signs, Symptoms or Conditions	A071205	Pacemaker Found in Back-Up Mode		
							F26	No Health Consequences or Impact	A071205	Pacemaker Found in Back-Up Mode		
					CT2406	Implantable	E2403	No Clinical Signs, Symptoms or Conditions	A071205	Pacemaker Found in Back-Up Mode		
							F26	No Health Consequences or Impact	A071205	Pacemaker Found in Back-Up Mode		
					CT2407	Active implantable	E2403	No Clinical Signs, Symptoms or Conditions	A071205	Pacemaker Found in Back-Up Mode		
							F26	No Health Consequences or Impact	A071205	Pacemaker Found in Back-Up Mode		
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A071205	Pacemaker Found in Back-Up Mode		
							F26	No Health Consequences or Impact	A071205	Pacemaker Found in Back-Up Mode		
					CT2727	Active implant devices/ancillaries	E2403	No Clinical Signs, Symptoms or Conditions	A071205	Pacemaker Found in Back-Up Mode		
							F26	No Health Consequences or Impact	A071205	Pacemaker Found in Back-Up Mode		
					CT2750	Implantable cardiac electrical stimulators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A071205	Pacemaker Found in Back-Up Mode		
							F26	No Health Consequences or Impact	A071205	Pacemaker Found in Back-Up Mode		
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A071205	Pacemaker Found in Back-Up Mode		
							F26	No Health Consequences or Impact	A071205	Pacemaker Found in Back-Up Mode		
					CT2985	Active	E2403	No Clinical Signs, Symptoms or Conditions	A071205	Pacemaker Found in Back-Up Mode		
							F26	No Health Consequences or Impact	A071205	Pacemaker Found in Back-Up Mode		
					CT3001	Implantable metal	E2403	No Clinical Signs, Symptoms or Conditions	A071205	Pacemaker Found in Back-Up Mode		
							F26	No Health Consequences or Impact	A071205	Pacemaker Found in Back-Up Mode		
					CT3002	Implantable synthetic polymer	E2403	No Clinical Signs, Symptoms or Conditions	A071205	Pacemaker Found in Back-Up Mode		
							F26	No Health Consequences or Impact	A071205	Pacemaker Found in Back-Up Mode		
					CT3006	Implantable inorganic material	E2403	No Clinical Signs, Symptoms or Conditions	A071205	Pacemaker Found in Back-Up Mode		
							F26	No Health Consequences or Impact	A071205	Pacemaker Found in Back-Up Mode		
					CT314	Electrically powered	E2403	No Clinical Signs, Symptoms or Conditions	A071205	Pacemaker Found in Back-Up Mode		
							F26	No Health Consequences or Impact	A071205	Pacemaker Found in Back-Up Mode		
					CT321	Long-term surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A071205	Pacemaker Found in Back-Up Mode		
							F26	No Health Consequences or Impact	A071205	Pacemaker Found in Back-Up Mode		
					CT449	Generators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A071205	Pacemaker Found in Back-Up Mode		
							F26	No Health Consequences or Impact	A071205	Pacemaker Found in Back-Up Mode		
					CT981	Single use	E2403	No Clinical Signs, Symptoms or Conditions	A071205	Pacemaker Found in Back-Up Mode		
							F26	No Health Consequences or Impact	A071205	Pacemaker Found in Back-Up Mode		
					CT983	Surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A071205	Pacemaker Found in Back-Up Mode		
							F26	No Health Consequences or Impact	A071205	Pacemaker Found in Back-Up Mode		
11695	2022	All other reportable incidents	47265	Dual-chamber implantable pacemaker, rate-responsive	CT141	Electrodes/Leads and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A071102	Failure to Interrogate	30/11/2022	30/11/2022
							F1901	Additional Surgery	A071102	Failure to Interrogate	30/11/2022	30/11/2022
					CT145	Cardiovascular devices	E2403	No Clinical Signs, Symptoms or Conditions	A071102	Failure to Interrogate	30/11/2022	30/11/2022
							F1901	Additional Surgery	A071102	Failure to Interrogate	30/11/2022	30/11/2022
					CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A071102	Failure to Interrogate	30/11/2022	30/11/2022
							F1901	Additional Surgery	A071102	Failure to Interrogate	30/11/2022	30/11/2022
					CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A071102	Failure to Interrogate	30/11/2022	30/11/2022
							F1901	Additional Surgery	A071102	Failure to Interrogate	30/11/2022	30/11/2022
					CT233	Surgical	E2403	No Clinical Signs, Symptoms or Conditions	A071102	Failure to Interrogate	30/11/2022	30/11/2022
							F1901	Additional Surgery	A071102	Failure to Interrogate	30/11/2022	30/11/2022
					CT2406	Implantable	E2403	No Clinical Signs, Symptoms or Conditions	A071102	Failure to Interrogate	30/11/2022	30/11/2022
							F1901	Additional Surgery	A071102	Failure to Interrogate	30/11/2022	30/11/2022
					CT2407	Active implantable	E2403	No Clinical Signs, Symptoms or Conditions	A071102	Failure to Interrogate	30/11/2022	30/11/2022
							F1901	Additional Surgery	A071102	Failure to Interrogate	30/11/2022	30/11/2022
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A071102	Failure to Interrogate	30/11/2022	30/11/2022
							F1901	Additional Surgery	A071102	Failure to Interrogate	30/11/2022	30/11/2022
					CT2727	Active implant devices/ancillaries	E2403	No Clinical Signs, Symptoms or Conditions	A071102	Failure to Interrogate	30/11/2022	30/11/2022
							F1901	Additional Surgery	A071102	Failure to Interrogate	30/11/2022	30/11/2022
					CT2750	Implantable cardiac electrical stimulators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A071102	Failure to Interrogate	30/11/2022	30/11/2022

						F1901	Additional Surgery	A071102	Failure to Interrogate	30/11/2022	30/11/2022
				CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A071102	Failure to Interrogate	30/11/2022	30/11/2022
						F1901	Additional Surgery	A071102	Failure to Interrogate	30/11/2022	30/11/2022
				CT2985	Active	E2403	No Clinical Signs, Symptoms or Conditions	A071102	Failure to Interrogate	30/11/2022	30/11/2022
						F1901	Additional Surgery	A071102	Failure to Interrogate	30/11/2022	30/11/2022
				CT3001	Implantable metal	E2403	No Clinical Signs, Symptoms or Conditions	A071102	Failure to Interrogate	30/11/2022	30/11/2022
						F1901	Additional Surgery	A071102	Failure to Interrogate	30/11/2022	30/11/2022
				CT3002	Implantable synthetic polymer	E2403	No Clinical Signs, Symptoms or Conditions	A071102	Failure to Interrogate	30/11/2022	30/11/2022
						F1901	Additional Surgery	A071102	Failure to Interrogate	30/11/2022	30/11/2022
				CT3006	Implantable inorganic material	E2403	No Clinical Signs, Symptoms or Conditions	A071102	Failure to Interrogate	30/11/2022	30/11/2022
						F1901	Additional Surgery	A071102	Failure to Interrogate	30/11/2022	30/11/2022
				CT314	Electrically powered	E2403	No Clinical Signs, Symptoms or Conditions	A071102	Failure to Interrogate	30/11/2022	30/11/2022
						F1901	Additional Surgery	A071102	Failure to Interrogate	30/11/2022	30/11/2022
				CT321	Long-term surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A071102	Failure to Interrogate	30/11/2022	30/11/2022
						F1901	Additional Surgery	A071102	Failure to Interrogate	30/11/2022	30/11/2022
				CT449	Generators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A071102	Failure to Interrogate	30/11/2022	30/11/2022
						F1901	Additional Surgery	A071102	Failure to Interrogate	30/11/2022	30/11/2022
				CT981	Single use	E2403	No Clinical Signs, Symptoms or Conditions	A071102	Failure to Interrogate	30/11/2022	30/11/2022
						F1901	Additional Surgery	A071102	Failure to Interrogate	30/11/2022	30/11/2022
				CT983	Surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A071102	Failure to Interrogate	30/11/2022	30/11/2022
						F1901	Additional Surgery	A071102	Failure to Interrogate	30/11/2022	30/11/2022
11696	2022	All other reportable incidents	37265	Dual-chamber implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	24/06/2022	24/06/2022
							F1901	Additional Surgery	A070504	Premature Discharge of Battery	24/06/2022	24/06/2022
11697	2022		47265	Dual-chamber implantable pacemaker, rate-responsive	CT141	Electrodes/Leads and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	22/11/2022	22/11/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	22/11/2022	22/11/2022
					CT145	Cardiovascular devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	22/11/2022	22/11/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	22/11/2022	22/11/2022
					CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	22/11/2022	22/11/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	22/11/2022	22/11/2022
					CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	22/11/2022	22/11/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	22/11/2022	22/11/2022
					CT233	Surgical	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	22/11/2022	22/11/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	22/11/2022	22/11/2022
					CT2406	Implantable	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	22/11/2022	22/11/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	22/11/2022	22/11/2022
					CT2407	Active implantable	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	22/11/2022	22/11/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	22/11/2022	22/11/2022
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	22/11/2022	22/11/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	22/11/2022	22/11/2022
					CT2727	Active implant devices/ancillaries	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	22/11/2022	22/11/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	22/11/2022	22/11/2022
					CT2750	Implantable cardiac electrical stimulators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	22/11/2022	22/11/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	22/11/2022	22/11/2022
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	22/11/2022	22/11/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	22/11/2022	22/11/2022
					CT2985	Active	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	22/11/2022	22/11/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	22/11/2022	22/11/2022
					CT3001	Implantable metal	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	22/11/2022	22/11/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	22/11/2022	22/11/2022
					CT3002	Implantable synthetic polymer	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	22/11/2022	22/11/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	22/11/2022	22/11/2022
					CT3006	Implantable inorganic material	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	22/11/2022	22/11/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	22/11/2022	22/11/2022
					CT314	Electrically powered	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	22/11/2022	22/11/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	22/11/2022	22/11/2022
					CT321	Long-term surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	22/11/2022	22/11/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	22/11/2022	22/11/2022
					CT449	Generators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	22/11/2022	22/11/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	22/11/2022	22/11/2022
					CT981	Single use	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	22/11/2022	22/11/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	22/11/2022	22/11/2022
					CT983	Surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	22/11/2022	22/11/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	22/11/2022	22/11/2022
11698	2022	All other reportable incidents	35223	Endocardial/interventricular septal pacing lead	CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070103	Intermittent Capture	10/11/2022	10/11/2022
									A0709	Device Sensing Problem	10/11/2022	10/11/2022
									A070909	Oversensing	10/11/2022	10/11/2022
									A070910	Undersensing	10/11/2022	10/11/2022
									A0722	Impedance Problem	10/11/2022	10/11/2022
							F1901	Additional Surgery	A070103	Intermittent Capture	10/11/2022	10/11/2022
									A0709	Device Sensing Problem	10/11/2022	10/11/2022
									A070909	Oversensing	10/11/2022	10/11/2022
									A070910	Undersensing	10/11/2022	10/11/2022
									A0722	Impedance Problem	10/11/2022	10/11/2022
11699	2022	All other reportable incidents	17882	Physiologic-monitoring defibrillation system	CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A26	Insufficient Device Problem Information		
							F27	Problem Identified During Non-Clinical Procedure	A26	Insufficient Device Problem Information		
11700	2022	All other reportable incidents	35223	Endocardial/interventricular septal pacing lead	CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070101	Failure to Capture		
									A070909	Oversensing		
									A072201	High Impedance		
							F1901	Additional Surgery	A070101	Failure to Capture		
									A070909	Oversensing		
									A072201	High Impedance		
11701	2022		47265	Dual-chamber implantable pacemaker, rate-responsive	CT141	Electrodes/Leads and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	21/09/2022	21/09/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	21/09/2022	21/09/2022
					CT145	Cardiovascular devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	21/09/2022	21/09/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	21/09/2022	21/09/2022
					CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	21/09/2022	21/09/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	21/09/2022	21/09/2022
					CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	21/09/2022	21/09/2022

							F1901	Additional Surgery	A24	No Apparent Device Problem	21/09/2022	21/09/2022
					CT233	Surgical	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	21/09/2022	21/09/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	21/09/2022	21/09/2022
					CT2406	Implantable	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	21/09/2022	21/09/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	21/09/2022	21/09/2022
					CT2407	Active implantable	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	21/09/2022	21/09/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	21/09/2022	21/09/2022
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	21/09/2022	21/09/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	21/09/2022	21/09/2022
					CT2727	Active implant devices/ancillaries	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	21/09/2022	21/09/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	21/09/2022	21/09/2022
					CT2750	Implantable cardiac electrical stimulators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	21/09/2022	21/09/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	21/09/2022	21/09/2022
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	21/09/2022	21/09/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	21/09/2022	21/09/2022
					CT2985	Active	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	21/09/2022	21/09/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	21/09/2022	21/09/2022
					CT3001	Implantable metal	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	21/09/2022	21/09/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	21/09/2022	21/09/2022
					CT3002	Implantable synthetic polymer	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	21/09/2022	21/09/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	21/09/2022	21/09/2022
					CT3006	Implantable inorganic material	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	21/09/2022	21/09/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	21/09/2022	21/09/2022
					CT314	Electrically powered	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	21/09/2022	21/09/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	21/09/2022	21/09/2022
					CT321	Long-term surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	21/09/2022	21/09/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	21/09/2022	21/09/2022
					CT449	Generators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	21/09/2022	21/09/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	21/09/2022	21/09/2022
					CT981	Single use	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	21/09/2022	21/09/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	21/09/2022	21/09/2022
					CT983	Surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	21/09/2022	21/09/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	21/09/2022	21/09/2022
11702	2022	All other reportable incidents	36241	Pacemaker/defibrillator lead	CT1684	Cardiac pacemakers and associated devices	E1906	Unspecified Infection	A24	No Apparent Device Problem		
							F08	Hospitalization or Prolonged Hospitalization	A24	No Apparent Device Problem		
							F1903	Device Explantation	A24	No Apparent Device Problem		
							F1906	Modified Surgical Procedure	A24	No Apparent Device Problem		
11703	2022		47265	Dual-chamber implantable pacemaker, rate-responsive	CT141	Electrodes/Leads and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	18/09/2022	18/09/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	18/09/2022	18/09/2022
					CT145	Cardiovascular devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	18/09/2022	18/09/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	18/09/2022	18/09/2022
					CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	18/09/2022	18/09/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	18/09/2022	18/09/2022
					CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	18/09/2022	18/09/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	18/09/2022	18/09/2022
					CT233	Surgical	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	18/09/2022	18/09/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	18/09/2022	18/09/2022
					CT2406	Implantable	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	18/09/2022	18/09/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	18/09/2022	18/09/2022
					CT2407	Active implantable	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	18/09/2022	18/09/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	18/09/2022	18/09/2022
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	18/09/2022	18/09/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	18/09/2022	18/09/2022
					CT2727	Active implant devices/ancillaries	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	18/09/2022	18/09/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	18/09/2022	18/09/2022
					CT2750	Implantable cardiac electrical stimulators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	18/09/2022	18/09/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	18/09/2022	18/09/2022
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	18/09/2022	18/09/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	18/09/2022	18/09/2022
					CT2985	Active	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	18/09/2022	18/09/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	18/09/2022	18/09/2022
					CT3001	Implantable metal	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	18/09/2022	18/09/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	18/09/2022	18/09/2022
					CT3002	Implantable synthetic polymer	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	18/09/2022	18/09/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	18/09/2022	18/09/2022
					CT3006	Implantable inorganic material	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	18/09/2022	18/09/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	18/09/2022	18/09/2022
					CT314	Electrically powered	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	18/09/2022	18/09/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	18/09/2022	18/09/2022
					CT321	Long-term surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	18/09/2022	18/09/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	18/09/2022	18/09/2022
					CT449	Generators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	18/09/2022	18/09/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	18/09/2022	18/09/2022
					CT981	Single use	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	18/09/2022	18/09/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	18/09/2022	18/09/2022
					CT983	Surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	18/09/2022	18/09/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	18/09/2022	18/09/2022
11704	2022		47265	Dual-chamber implantable pacemaker, rate-responsive	CT141	Electrodes/Leads and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	02/12/2022	02/12/2022
							F1901	Additional Surgery	A070504	Premature Discharge of Battery	02/12/2022	02/12/2022
					CT145	Cardiovascular devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	02/12/2022	02/12/2022
							F1901	Additional Surgery	A070504	Premature Discharge of Battery	02/12/2022	02/12/2022
					CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	02/12/2022	02/12/2022
							F1901	Additional Surgery	A070504	Premature Discharge of Battery	02/12/2022	02/12/2022
					CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	02/12/2022	02/12/2022
							F1901	Additional Surgery	A070504	Premature Discharge of Battery	02/12/2022	02/12/2022
					CT233	Surgical	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	02/12/2022	02/12/2022
							F1901	Additional Surgery	A070504	Premature Discharge of Battery	02/12/2022	02/12/2022
					CT2406	Implantable	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	02/12/2022	02/12/2022

							F1901	Additional Surgery	A070504	Premature Discharge of Battery	02/12/2022	02/12/2022
					CT2407	Active implantable	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	02/12/2022	02/12/2022
							F1901	Additional Surgery	A070504	Premature Discharge of Battery	02/12/2022	02/12/2022
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	02/12/2022	02/12/2022
							F1901	Additional Surgery	A070504	Premature Discharge of Battery	02/12/2022	02/12/2022
					CT2727	Active implant devices/ancillaries	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	02/12/2022	02/12/2022
							F1901	Additional Surgery	A070504	Premature Discharge of Battery	02/12/2022	02/12/2022
					CT2750	Implantable cardiac electrical stimulators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	02/12/2022	02/12/2022
							F1901	Additional Surgery	A070504	Premature Discharge of Battery	02/12/2022	02/12/2022
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	02/12/2022	02/12/2022
							F1901	Additional Surgery	A070504	Premature Discharge of Battery	02/12/2022	02/12/2022
					CT2985	Active	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	02/12/2022	02/12/2022
							F1901	Additional Surgery	A070504	Premature Discharge of Battery	02/12/2022	02/12/2022
					CT3001	Implantable metal	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	02/12/2022	02/12/2022
							F1901	Additional Surgery	A070504	Premature Discharge of Battery	02/12/2022	02/12/2022
					CT3002	Implantable synthetic polymer	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	02/12/2022	02/12/2022
							F1901	Additional Surgery	A070504	Premature Discharge of Battery	02/12/2022	02/12/2022
					CT3006	Implantable inorganic material	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	02/12/2022	02/12/2022
							F1901	Additional Surgery	A070504	Premature Discharge of Battery	02/12/2022	02/12/2022
					CT314	Electrically powered	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	02/12/2022	02/12/2022
							F1901	Additional Surgery	A070504	Premature Discharge of Battery	02/12/2022	02/12/2022
					CT321	Long-term surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	02/12/2022	02/12/2022
							F1901	Additional Surgery	A070504	Premature Discharge of Battery	02/12/2022	02/12/2022
					CT449	Generators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	02/12/2022	02/12/2022
							F1901	Additional Surgery	A070504	Premature Discharge of Battery	02/12/2022	02/12/2022
					CT981	Single use	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	02/12/2022	02/12/2022
							F1901	Additional Surgery	A070504	Premature Discharge of Battery	02/12/2022	02/12/2022
					CT983	Surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	02/12/2022	02/12/2022
							F1901	Additional Surgery	A070504	Premature Discharge of Battery	02/12/2022	02/12/2022
11705	2022	All other reportable incidents	47263	Cardiac resynchronization therapy implantable pacemaker	CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	27/10/2022	27/10/2022
							F08	Hospitalization or Prolonged Hospitalization	A24	No Apparent Device Problem	27/10/2022	27/10/2022
							F1905	Device Revision or Replacement	A24	No Apparent Device Problem	27/10/2022	27/10/2022
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	27/10/2022	27/10/2022
							F08	Hospitalization or Prolonged Hospitalization	A24	No Apparent Device Problem	27/10/2022	27/10/2022
							F1905	Device Revision or Replacement	A24	No Apparent Device Problem	27/10/2022	27/10/2022
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	27/10/2022	27/10/2022
							F08	Hospitalization or Prolonged Hospitalization	A24	No Apparent Device Problem	27/10/2022	27/10/2022
							F1905	Device Revision or Replacement	A24	No Apparent Device Problem	27/10/2022	27/10/2022
11706	2022	All other reportable incidents	47263	Cardiac resynchronization therapy implantable pacemaker	CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A072201	High Impedance		
									A090801	Signal Artifact/Noise		
							F26	No Health Consequences or Impact	A072201	High Impedance		
									A090801	Signal Artifact/Noise		
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A072201	High Impedance		
									A090801	Signal Artifact/Noise		
							F26	No Health Consequences or Impact	A072201	High Impedance		
									A090801	Signal Artifact/Noise		
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A072201	High Impedance		
									A090801	Signal Artifact/Noise		
							F26	No Health Consequences or Impact	A072201	High Impedance		
									A090801	Signal Artifact/Noise		
11707	2022	All other reportable incidents	37265	Dual-chamber implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F24	Insufficient Health Impact Information	A070504	Premature Discharge of Battery		
11708	2022	All other reportable incidents	17882	Physiologic-monitoring defibrillation system	CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A1603	Failure of Device to Self-Test		
							F27	Problem Identified During Non-Clinical Procedure	A1603	Failure of Device to Self-Test		
11709	2022		47265	Dual-chamber implantable pacemaker, rate-responsive	CT141	Electrodes/Leads and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	01/12/2022	01/12/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	01/12/2022	01/12/2022
					CT145	Cardiovascular devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	01/12/2022	01/12/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	01/12/2022	01/12/2022
					CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	01/12/2022	01/12/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	01/12/2022	01/12/2022
					CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	01/12/2022	01/12/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	01/12/2022	01/12/2022
					CT233	Surgical	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	01/12/2022	01/12/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	01/12/2022	01/12/2022
					CT2406	Implantable	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	01/12/2022	01/12/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	01/12/2022	01/12/2022
					CT2407	Active implantable	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	01/12/2022	01/12/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	01/12/2022	01/12/2022
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	01/12/2022	01/12/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	01/12/2022	01/12/2022
					CT2727	Active implant devices/ancillaries	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	01/12/2022	01/12/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	01/12/2022	01/12/2022
					CT2750	Implantable cardiac electrical stimulators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	01/12/2022	01/12/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	01/12/2022	01/12/2022
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	01/12/2022	01/12/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	01/12/2022	01/12/2022
					CT2985	Active	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	01/12/2022	01/12/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	01/12/2022	01/12/2022
					CT3001	Implantable metal	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	01/12/2022	01/12/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	01/12/2022	01/12/2022
					CT3002	Implantable synthetic polymer	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	01/12/2022	01/12/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	01/12/2022	01/12/2022
					CT3006	Implantable inorganic material	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	01/12/2022	01/12/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	01/12/2022	01/12/2022
					CT314	Electrically powered	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	01/12/2022	01/12/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	01/12/2022	01/12/2022
					CT321	Long-term surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	01/12/2022	01/12/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	01/12/2022	01/12/2022

					CT449	Generators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	01/12/2022	01/12/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	01/12/2022	01/12/2022
					CT981	Single use	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	01/12/2022	01/12/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	01/12/2022	01/12/2022
					CT983	Surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	01/12/2022	01/12/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	01/12/2022	01/12/2022
11710	2022		47265	Dual-chamber implantable pacemaker, rate-responsive	CT141	Electrodes/Leads and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	05/10/2022	05/10/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	05/10/2022	05/10/2022
					CT145	Cardiovascular devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	05/10/2022	05/10/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	05/10/2022	05/10/2022
					CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	05/10/2022	05/10/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	05/10/2022	05/10/2022
					CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	05/10/2022	05/10/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	05/10/2022	05/10/2022
					CT233	Surgical	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	05/10/2022	05/10/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	05/10/2022	05/10/2022
					CT2406	Implantable	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	05/10/2022	05/10/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	05/10/2022	05/10/2022
					CT2407	Active implantable	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	05/10/2022	05/10/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	05/10/2022	05/10/2022
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	05/10/2022	05/10/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	05/10/2022	05/10/2022
					CT2727	Active implant devices/ancillaries	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	05/10/2022	05/10/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	05/10/2022	05/10/2022
					CT2750	Implantable cardiac electrical stimulators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	05/10/2022	05/10/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	05/10/2022	05/10/2022
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	05/10/2022	05/10/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	05/10/2022	05/10/2022
					CT2985	Active	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	05/10/2022	05/10/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	05/10/2022	05/10/2022
					CT3001	Implantable metal	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	05/10/2022	05/10/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	05/10/2022	05/10/2022
					CT3002	Implantable synthetic polymer	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	05/10/2022	05/10/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	05/10/2022	05/10/2022
					CT3006	Implantable inorganic material	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	05/10/2022	05/10/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	05/10/2022	05/10/2022
					CT314	Electrically powered	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	05/10/2022	05/10/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	05/10/2022	05/10/2022
					CT321	Long-term surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	05/10/2022	05/10/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	05/10/2022	05/10/2022
					CT449	Generators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	05/10/2022	05/10/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	05/10/2022	05/10/2022
					CT981	Single use	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	05/10/2022	05/10/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	05/10/2022	05/10/2022
					CT983	Surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	05/10/2022	05/10/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	05/10/2022	05/10/2022
11711	2022		47265	Dual-chamber implantable pacemaker, rate-responsive	CT141	Electrodes/Leads and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	26/10/2022	26/10/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	26/10/2022	26/10/2022
					CT145	Cardiovascular devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	26/10/2022	26/10/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	26/10/2022	26/10/2022
					CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	26/10/2022	26/10/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	26/10/2022	26/10/2022
					CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	26/10/2022	26/10/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	26/10/2022	26/10/2022
					CT233	Surgical	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	26/10/2022	26/10/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	26/10/2022	26/10/2022
					CT2406	Implantable	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	26/10/2022	26/10/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	26/10/2022	26/10/2022
					CT2407	Active implantable	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	26/10/2022	26/10/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	26/10/2022	26/10/2022
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	26/10/2022	26/10/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	26/10/2022	26/10/2022
					CT2727	Active implant devices/ancillaries	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	26/10/2022	26/10/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	26/10/2022	26/10/2022
					CT2750	Implantable cardiac electrical stimulators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	26/10/2022	26/10/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	26/10/2022	26/10/2022
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	26/10/2022	26/10/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	26/10/2022	26/10/2022
					CT2985	Active	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	26/10/2022	26/10/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	26/10/2022	26/10/2022
					CT3001	Implantable metal	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	26/10/2022	26/10/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	26/10/2022	26/10/2022
					CT3002	Implantable synthetic polymer	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	26/10/2022	26/10/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	26/10/2022	26/10/2022
					CT3006	Implantable inorganic material	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	26/10/2022	26/10/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	26/10/2022	26/10/2022
					CT314	Electrically powered	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	26/10/2022	26/10/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	26/10/2022	26/10/2022
					CT321	Long-term surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	26/10/2022	26/10/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	26/10/2022	26/10/2022
					CT449	Generators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	26/10/2022	26/10/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	26/10/2022	26/10/2022
					CT981	Single use	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	26/10/2022	26/10/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	26/10/2022	26/10/2022
					CT983	Surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	26/10/2022	26/10/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	26/10/2022	26/10/2022
11712	2022		47265	Dual-chamber implantable pacemaker, rate-responsive	CT141	Electrodes/Leads and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	31/10/2022	31/10/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	31/10/2022	31/10/2022

					CT145	Cardiovascular devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	31/10/2022	31/10/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	31/10/2022	31/10/2022
					CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	31/10/2022	31/10/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	31/10/2022	31/10/2022
					CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	31/10/2022	31/10/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	31/10/2022	31/10/2022
					CT233	Surgical	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	31/10/2022	31/10/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	31/10/2022	31/10/2022
					CT2406	Implantable	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	31/10/2022	31/10/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	31/10/2022	31/10/2022
					CT2407	Active implantable	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	31/10/2022	31/10/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	31/10/2022	31/10/2022
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	31/10/2022	31/10/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	31/10/2022	31/10/2022
					CT2727	Active implant devices/ancillaries	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	31/10/2022	31/10/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	31/10/2022	31/10/2022
					CT2750	Implantable cardiac electrical stimulators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	31/10/2022	31/10/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	31/10/2022	31/10/2022
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	31/10/2022	31/10/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	31/10/2022	31/10/2022
					CT2985	Active	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	31/10/2022	31/10/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	31/10/2022	31/10/2022
					CT3001	Implantable metal	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	31/10/2022	31/10/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	31/10/2022	31/10/2022
					CT3002	Implantable synthetic polymer	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	31/10/2022	31/10/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	31/10/2022	31/10/2022
					CT3006	Implantable inorganic material	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	31/10/2022	31/10/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	31/10/2022	31/10/2022
					CT314	Electrically powered	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	31/10/2022	31/10/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	31/10/2022	31/10/2022
					CT321	Long-term surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	31/10/2022	31/10/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	31/10/2022	31/10/2022
					CT449	Generators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	31/10/2022	31/10/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	31/10/2022	31/10/2022
					CT981	Single use	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	31/10/2022	31/10/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	31/10/2022	31/10/2022
					CT983	Surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	31/10/2022	31/10/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	31/10/2022	31/10/2022
11713	2022		47265	Dual-chamber implantable pacemaker, rate-responsive	CT141	Electrodes/Leads and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	21/11/2022	21/11/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	21/11/2022	21/11/2022
					CT145	Cardiovascular devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	21/11/2022	21/11/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	21/11/2022	21/11/2022
					CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	21/11/2022	21/11/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	21/11/2022	21/11/2022
					CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	21/11/2022	21/11/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	21/11/2022	21/11/2022
					CT233	Surgical	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	21/11/2022	21/11/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	21/11/2022	21/11/2022
					CT2406	Implantable	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	21/11/2022	21/11/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	21/11/2022	21/11/2022
					CT2407	Active implantable	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	21/11/2022	21/11/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	21/11/2022	21/11/2022
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	21/11/2022	21/11/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	21/11/2022	21/11/2022
					CT2727	Active implant devices/ancillaries	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	21/11/2022	21/11/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	21/11/2022	21/11/2022
					CT2750	Implantable cardiac electrical stimulators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	21/11/2022	21/11/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	21/11/2022	21/11/2022
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	21/11/2022	21/11/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	21/11/2022	21/11/2022
					CT2985	Active	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	21/11/2022	21/11/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	21/11/2022	21/11/2022
					CT3001	Implantable metal	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	21/11/2022	21/11/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	21/11/2022	21/11/2022
					CT3002	Implantable synthetic polymer	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	21/11/2022	21/11/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	21/11/2022	21/11/2022
					CT3006	Implantable inorganic material	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	21/11/2022	21/11/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	21/11/2022	21/11/2022
					CT314	Electrically powered	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	21/11/2022	21/11/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	21/11/2022	21/11/2022
					CT321	Long-term surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	21/11/2022	21/11/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	21/11/2022	21/11/2022
					CT449	Generators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	21/11/2022	21/11/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	21/11/2022	21/11/2022
					CT981	Single use	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	21/11/2022	21/11/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	21/11/2022	21/11/2022
					CT983	Surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	21/11/2022	21/11/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	21/11/2022	21/11/2022
11714	2022		47265	Dual-chamber implantable pacemaker, rate-responsive	CT141	Electrodes/Leads and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	02/11/2022	02/11/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	02/11/2022	02/11/2022
					CT145	Cardiovascular devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	02/11/2022	02/11/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	02/11/2022	02/11/2022
					CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	02/11/2022	02/11/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	02/11/2022	02/11/2022
					CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	02/11/2022	02/11/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	02/11/2022	02/11/2022
					CT233	Surgical	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	02/11/2022	02/11/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	02/11/2022	02/11/2022

					CT2406	Implantable	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	02/11/2022	02/11/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	02/11/2022	02/11/2022
					CT2407	Active implantable	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	02/11/2022	02/11/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	02/11/2022	02/11/2022
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	02/11/2022	02/11/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	02/11/2022	02/11/2022
					CT2727	Active implant devices/ancillaries	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	02/11/2022	02/11/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	02/11/2022	02/11/2022
					CT2750	Implantable cardiac electrical stimulators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	02/11/2022	02/11/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	02/11/2022	02/11/2022
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	02/11/2022	02/11/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	02/11/2022	02/11/2022
					CT2985	Active	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	02/11/2022	02/11/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	02/11/2022	02/11/2022
					CT3001	Implantable metal	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	02/11/2022	02/11/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	02/11/2022	02/11/2022
					CT3002	Implantable synthetic polymer	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	02/11/2022	02/11/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	02/11/2022	02/11/2022
					CT3006	Implantable inorganic material	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	02/11/2022	02/11/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	02/11/2022	02/11/2022
					CT314	Electrically powered	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	02/11/2022	02/11/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	02/11/2022	02/11/2022
					CT321	Long-term surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	02/11/2022	02/11/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	02/11/2022	02/11/2022
					CT449	Generators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	02/11/2022	02/11/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	02/11/2022	02/11/2022
					CT981	Single use	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	02/11/2022	02/11/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	02/11/2022	02/11/2022
					CT983	Surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	02/11/2022	02/11/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	02/11/2022	02/11/2022
11715	2022		47265	Dual-chamber implantable pacemaker, rate-responsive	CT141	Electrodes/Leads and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	09/11/2022	09/11/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	09/11/2022	09/11/2022
					CT145	Cardiovascular devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	09/11/2022	09/11/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	09/11/2022	09/11/2022
					CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	09/11/2022	09/11/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	09/11/2022	09/11/2022
					CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	09/11/2022	09/11/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	09/11/2022	09/11/2022
					CT233	Surgical	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	09/11/2022	09/11/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	09/11/2022	09/11/2022
					CT2406	Implantable	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	09/11/2022	09/11/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	09/11/2022	09/11/2022
					CT2407	Active implantable	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	09/11/2022	09/11/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	09/11/2022	09/11/2022
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	09/11/2022	09/11/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	09/11/2022	09/11/2022
					CT2727	Active implant devices/ancillaries	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	09/11/2022	09/11/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	09/11/2022	09/11/2022
					CT2750	Implantable cardiac electrical stimulators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	09/11/2022	09/11/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	09/11/2022	09/11/2022
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	09/11/2022	09/11/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	09/11/2022	09/11/2022
					CT2985	Active	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	09/11/2022	09/11/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	09/11/2022	09/11/2022
					CT3001	Implantable metal	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	09/11/2022	09/11/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	09/11/2022	09/11/2022
					CT3002	Implantable synthetic polymer	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	09/11/2022	09/11/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	09/11/2022	09/11/2022
					CT3006	Implantable inorganic material	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	09/11/2022	09/11/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	09/11/2022	09/11/2022
					CT314	Electrically powered	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	09/11/2022	09/11/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	09/11/2022	09/11/2022
					CT321	Long-term surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	09/11/2022	09/11/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	09/11/2022	09/11/2022
					CT449	Generators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	09/11/2022	09/11/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	09/11/2022	09/11/2022
					CT981	Single use	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	09/11/2022	09/11/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	09/11/2022	09/11/2022
					CT983	Surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	09/11/2022	09/11/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	09/11/2022	09/11/2022
11716	2022	All other reportable incidents	35852	Single-chamber implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A1304	Telemetry Discrepancy		
							F26	No Health Consequences or Impact	A1304	Telemetry Discrepancy		
11717	2022		47265	Dual-chamber implantable pacemaker, rate-responsive	CT141	Electrodes/Leads and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	30/11/2022	30/11/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	30/11/2022	30/11/2022
					CT145	Cardiovascular devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	30/11/2022	30/11/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	30/11/2022	30/11/2022
					CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	30/11/2022	30/11/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	30/11/2022	30/11/2022
					CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	30/11/2022	30/11/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	30/11/2022	30/11/2022
					CT233	Surgical	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	30/11/2022	30/11/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	30/11/2022	30/11/2022
					CT2406	Implantable	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	30/11/2022	30/11/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	30/11/2022	30/11/2022
					CT2407	Active implantable	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	30/11/2022	30/11/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	30/11/2022	30/11/2022
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	30/11/2022	30/11/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	30/11/2022	30/11/2022

					CT2727	Active implant devices/ancillaries	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	30/11/2022	30/11/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	30/11/2022	30/11/2022
					CT2750	Implantable cardiac electrical stimulators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	30/11/2022	30/11/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	30/11/2022	30/11/2022
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	30/11/2022	30/11/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	30/11/2022	30/11/2022
					CT2985	Active	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	30/11/2022	30/11/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	30/11/2022	30/11/2022
					CT3001	Implantable metal	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	30/11/2022	30/11/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	30/11/2022	30/11/2022
					CT3002	Implantable synthetic polymer	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	30/11/2022	30/11/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	30/11/2022	30/11/2022
					CT3006	Implantable inorganic material	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	30/11/2022	30/11/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	30/11/2022	30/11/2022
					CT314	Electrically powered	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	30/11/2022	30/11/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	30/11/2022	30/11/2022
					CT321	Long-term surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	30/11/2022	30/11/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	30/11/2022	30/11/2022
					CT449	Generators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	30/11/2022	30/11/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	30/11/2022	30/11/2022
					CT981	Single use	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	30/11/2022	30/11/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	30/11/2022	30/11/2022
					CT983	Surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	30/11/2022	30/11/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	30/11/2022	30/11/2022
11718	2022	All other reportable incidents	47265	Dual-chamber implantable pacemaker, rate-responsive	CT141	Electrodes/Leads and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	22/11/2022	22/11/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	22/11/2022	22/11/2022
					CT145	Cardiovascular devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	22/11/2022	22/11/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	22/11/2022	22/11/2022
					CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	22/11/2022	22/11/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	22/11/2022	22/11/2022
					CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	22/11/2022	22/11/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	22/11/2022	22/11/2022
					CT233	Surgical	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	22/11/2022	22/11/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	22/11/2022	22/11/2022
					CT2406	Implantable	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	22/11/2022	22/11/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	22/11/2022	22/11/2022
					CT2407	Active implantable	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	22/11/2022	22/11/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	22/11/2022	22/11/2022
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	22/11/2022	22/11/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	22/11/2022	22/11/2022
					CT2727	Active implant devices/ancillaries	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	22/11/2022	22/11/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	22/11/2022	22/11/2022
					CT2750	Implantable cardiac electrical stimulators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	22/11/2022	22/11/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	22/11/2022	22/11/2022
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	22/11/2022	22/11/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	22/11/2022	22/11/2022
					CT2985	Active	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	22/11/2022	22/11/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	22/11/2022	22/11/2022
					CT3001	Implantable metal	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	22/11/2022	22/11/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	22/11/2022	22/11/2022
					CT3002	Implantable synthetic polymer	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	22/11/2022	22/11/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	22/11/2022	22/11/2022
					CT3006	Implantable inorganic material	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	22/11/2022	22/11/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	22/11/2022	22/11/2022
					CT314	Electrically powered	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	22/11/2022	22/11/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	22/11/2022	22/11/2022
					CT321	Long-term surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	22/11/2022	22/11/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	22/11/2022	22/11/2022
					CT449	Generators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	22/11/2022	22/11/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	22/11/2022	22/11/2022
					CT981	Single use	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	22/11/2022	22/11/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	22/11/2022	22/11/2022
					CT983	Surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	22/11/2022	22/11/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	22/11/2022	22/11/2022
11719	2022	All other reportable incidents	47265	Dual-chamber implantable pacemaker, rate-responsive	CT141	Electrodes/Leads and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT145	Cardiovascular devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT233	Surgical	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT2406	Implantable	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT2407	Active implantable	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT2727	Active implant devices/ancillaries	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT2750	Implantable cardiac electrical stimulators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT2985	Active	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		

					CT3001	Implantable metal	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT3002	Implantable synthetic polymer	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT3006	Implantable inorganic material	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT314	Electrically powered	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT321	Long-term surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT449	Generators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT981	Single use	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT983	Surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
11720	2022	All other reportable incidents	60956	Pacing/defibrillation lead tunneller	CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A0203	Defective Device	29/09/2022	29/09/2022
							F24	Insufficient Health Impact Information	A0203	Defective Device	29/09/2022	29/09/2022
11721	2022	All other reportable incidents	47270	Cardiac resynchronization therapy implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A072201	High Impedance		
							F26	No Health Consequences or Impact	A072201	High Impedance		
11722	2022	All other reportable incidents	35223	Endocardial/interventricular septal pacing lead	CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A0401	Break	20/09/2022	20/09/2022
									A070909	Oversensing	20/09/2022	20/09/2022
									A0712	Pacing Problem	20/09/2022	20/09/2022
							F1901	Additional Surgery	A0401	Break	20/09/2022	20/09/2022
									A070909	Oversensing	20/09/2022	20/09/2022
									A0712	Pacing Problem	20/09/2022	20/09/2022
11723	2022	All other reportable incidents	47263	Cardiac resynchronization therapy implantable pacemaker	CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	23/02/2022	23/02/2022
									A0908	Incorrect, Inadequate or Imprecise Result or Readings	23/02/2022	23/02/2022
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	23/02/2022	23/02/2022
									A0908	Incorrect, Inadequate or Imprecise Result or Readings	23/02/2022	23/02/2022
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	23/02/2022	23/02/2022
									A0908	Incorrect, Inadequate or Imprecise Result or Readings	23/02/2022	23/02/2022
					CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	23/02/2022	23/02/2022
									A0908	Incorrect, Inadequate or Imprecise Result or Readings	23/02/2022	23/02/2022
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	23/02/2022	23/02/2022
									A0908	Incorrect, Inadequate or Imprecise Result or Readings	23/02/2022	23/02/2022
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	23/02/2022	23/02/2022
									A0908	Incorrect, Inadequate or Imprecise Result or Readings	23/02/2022	23/02/2022
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	23/02/2022	23/02/2022
									A0908	Incorrect, Inadequate or Imprecise Result or Readings	23/02/2022	23/02/2022
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	23/02/2022	23/02/2022
									A0908	Incorrect, Inadequate or Imprecise Result or Readings	23/02/2022	23/02/2022
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	23/02/2022	23/02/2022
									A0908	Incorrect, Inadequate or Imprecise Result or Readings	23/02/2022	23/02/2022
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	23/02/2022	23/02/2022
									A0908	Incorrect, Inadequate or Imprecise Result or Readings	23/02/2022	23/02/2022
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	23/02/2022	23/02/2022
									A0908	Incorrect, Inadequate or Imprecise Result or Readings	23/02/2022	23/02/2022
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	23/02/2022	23/02/2022
									A0908	Incorrect, Inadequate or Imprecise Result or Readings	23/02/2022	23/02/2022
11724	2022	All other reportable incidents	47263	Cardiac resynchronization therapy implantable pacemaker	CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	23/02/2022	23/02/2022
									A0908	Incorrect, Inadequate or Imprecise Result or Readings	23/02/2022	23/02/2022
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	23/02/2022	23/02/2022
									A0908	Incorrect, Inadequate or Imprecise Result or Readings	23/02/2022	23/02/2022
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	23/02/2022	23/02/2022
									A0908	Incorrect, Inadequate or Imprecise Result or Readings	23/02/2022	23/02/2022
					CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	23/02/2022	23/02/2022
									A0908	Incorrect, Inadequate or Imprecise Result or Readings	23/02/2022	23/02/2022
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	23/02/2022	23/02/2022
									A0908	Incorrect, Inadequate or Imprecise Result or Readings	23/02/2022	23/02/2022
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	23/02/2022	23/02/2022
									A0908	Incorrect, Inadequate or Imprecise Result or Readings	23/02/2022	23/02/2022
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	23/02/2022	23/02/2022
									A0908	Incorrect, Inadequate or Imprecise Result or Readings	23/02/2022	23/02/2022
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	23/02/2022	23/02/2022
									A0908	Incorrect, Inadequate or Imprecise Result or Readings	23/02/2022	23/02/2022
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	23/02/2022	23/02/2022
									A0908	Incorrect, Inadequate or Imprecise Result or Readings	23/02/2022	23/02/2022
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	23/02/2022	23/02/2022
									A0908	Incorrect, Inadequate or Imprecise Result or Readings	23/02/2022	23/02/2022
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	23/02/2022	23/02/2022

								A0908	Incorrect, Inadequate or Imprecise Result or Readings	23/02/2022	23/02/2022
						F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	23/02/2022	23/02/2022
								A0908	Incorrect, Inadequate or Imprecise Result or Readings	23/02/2022	23/02/2022
11725	2022		47265	Dual-chamber implantable pacemaker, rate-responsive	CT141	Electrodes/Leads and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	30/11/2022	30/11/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	30/11/2022	30/11/2022
					CT145	Cardiovascular devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	30/11/2022	30/11/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	30/11/2022	30/11/2022
					CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	30/11/2022	30/11/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	30/11/2022	30/11/2022
					CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	30/11/2022	30/11/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	30/11/2022	30/11/2022
					CT233	Surgical	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	30/11/2022	30/11/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	30/11/2022	30/11/2022
					CT2406	Implantable	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	30/11/2022	30/11/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	30/11/2022	30/11/2022
					CT2407	Active implantable	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	30/11/2022	30/11/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	30/11/2022	30/11/2022
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	30/11/2022	30/11/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	30/11/2022	30/11/2022
					CT2727	Active implant devices/ancillaries	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	30/11/2022	30/11/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	30/11/2022	30/11/2022
					CT2750	Implantable cardiac electrical stimulators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	30/11/2022	30/11/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	30/11/2022	30/11/2022
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	30/11/2022	30/11/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	30/11/2022	30/11/2022
					CT2985	Active	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	30/11/2022	30/11/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	30/11/2022	30/11/2022
					CT3001	Implantable metal	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	30/11/2022	30/11/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	30/11/2022	30/11/2022
					CT3002	Implantable synthetic polymer	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	30/11/2022	30/11/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	30/11/2022	30/11/2022
					CT3006	Implantable inorganic material	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	30/11/2022	30/11/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	30/11/2022	30/11/2022
					CT314	Electrically powered	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	30/11/2022	30/11/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	30/11/2022	30/11/2022
					CT321	Long-term surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	30/11/2022	30/11/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	30/11/2022	30/11/2022
					CT449	Generators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	30/11/2022	30/11/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	30/11/2022	30/11/2022
					CT981	Single use	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	30/11/2022	30/11/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	30/11/2022	30/11/2022
					CT983	Surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	30/11/2022	30/11/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	30/11/2022	30/11/2022
11726	2022	All other reportable incidents	35223	Endocardial/interventricular septal pacing lead	CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A0401	Break	30/06/2022	30/06/2022
									A070909	Oversensing	30/06/2022	30/06/2022
							F1901	Additional Surgery	A0401	Break	30/06/2022	30/06/2022
									A070909	Oversensing	30/06/2022	30/06/2022
11727	2022	All other reportable incidents	47265	Dual-chamber implantable pacemaker, rate-responsive	CT141	Electrodes/Leads and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem		
							F26	No Health Consequences or Impact	A24	No Apparent Device Problem		
					CT145	Cardiovascular devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem		
							F26	No Health Consequences or Impact	A24	No Apparent Device Problem		
					CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem		
							F26	No Health Consequences or Impact	A24	No Apparent Device Problem		
					CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem		
							F26	No Health Consequences or Impact	A24	No Apparent Device Problem		
					CT233	Surgical	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem		
							F26	No Health Consequences or Impact	A24	No Apparent Device Problem		
					CT2406	Implantable	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem		
							F26	No Health Consequences or Impact	A24	No Apparent Device Problem		
					CT2407	Active implantable	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem		
							F26	No Health Consequences or Impact	A24	No Apparent Device Problem		
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem		
							F26	No Health Consequences or Impact	A24	No Apparent Device Problem		
					CT2727	Active implant devices/ancillaries	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem		
							F26	No Health Consequences or Impact	A24	No Apparent Device Problem		
					CT2750	Implantable cardiac electrical stimulators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem		
							F26	No Health Consequences or Impact	A24	No Apparent Device Problem		
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem		
							F26	No Health Consequences or Impact	A24	No Apparent Device Problem		
					CT2985	Active	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem		
							F26	No Health Consequences or Impact	A24	No Apparent Device Problem		
					CT3001	Implantable metal	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem		
							F26	No Health Consequences or Impact	A24	No Apparent Device Problem		
					CT3002	Implantable synthetic polymer	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem		
							F26	No Health Consequences or Impact	A24	No Apparent Device Problem		
					CT3006	Implantable inorganic material	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem		
							F26	No Health Consequences or Impact	A24	No Apparent Device Problem		
					CT314	Electrically powered	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem		
							F26	No Health Consequences or Impact	A24	No Apparent Device Problem		
					CT321	Long-term surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem		
							F26	No Health Consequences or Impact	A24	No Apparent Device Problem		
					CT449	Generators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem		
							F26	No Health Consequences or Impact	A24	No Apparent Device Problem		
					CT981	Single use	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem		
							F26	No Health Consequences or Impact	A24	No Apparent Device Problem		
					CT983	Surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem		

							F26	No Health Consequences or Impact	A24	No Apparent Device Problem		
11728	2022	All other reportable incidents	47265	Dual-chamber implantable pacemaker, rate-responsive	CT141	Electrodes/Leads and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem		
							F1901	Additional Surgery	A24	No Apparent Device Problem		
					CT145	Cardiovascular devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem		
							F1901	Additional Surgery	A24	No Apparent Device Problem		
					CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem		
							F1901	Additional Surgery	A24	No Apparent Device Problem		
					CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem		
							F1901	Additional Surgery	A24	No Apparent Device Problem		
					CT233	Surgical	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem		
							F1901	Additional Surgery	A24	No Apparent Device Problem		
					CT2406	Implantable	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem		
							F1901	Additional Surgery	A24	No Apparent Device Problem		
					CT2407	Active implantable	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem		
							F1901	Additional Surgery	A24	No Apparent Device Problem		
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem		
							F1901	Additional Surgery	A24	No Apparent Device Problem		
					CT2727	Active implant devices/ancillaries	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem		
							F1901	Additional Surgery	A24	No Apparent Device Problem		
					CT2750	Implantable cardiac electrical stimulators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem		
							F1901	Additional Surgery	A24	No Apparent Device Problem		
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem		
							F1901	Additional Surgery	A24	No Apparent Device Problem		
					CT2985	Active	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem		
							F1901	Additional Surgery	A24	No Apparent Device Problem		
					CT3001	Implantable metal	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem		
							F1901	Additional Surgery	A24	No Apparent Device Problem		
					CT3002	Implantable synthetic polymer	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem		
							F1901	Additional Surgery	A24	No Apparent Device Problem		
					CT3006	Implantable inorganic material	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem		
							F1901	Additional Surgery	A24	No Apparent Device Problem		
					CT314	Electrically powered	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem		
							F1901	Additional Surgery	A24	No Apparent Device Problem		
					CT321	Long-term surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem		
							F1901	Additional Surgery	A24	No Apparent Device Problem		
					CT449	Generators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem		
							F1901	Additional Surgery	A24	No Apparent Device Problem		
					CT981	Single use	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem		
							F1901	Additional Surgery	A24	No Apparent Device Problem		
					CT983	Surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem		
							F1901	Additional Surgery	A24	No Apparent Device Problem		
11729	2022	All other reportable incidents	47270	Cardiac resynchronization therapy implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A072201	High Impedance		
							F26	No Health Consequences or Impact	A072201	High Impedance		
11730	2022	All other reportable incidents	47263	Cardiac resynchronization therapy implantable pacemaker	CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	01/07/2023	01/07/2023
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	01/07/2023	01/07/2023
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	01/07/2023	01/07/2023
					CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	01/07/2023	01/07/2023
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	01/07/2023	01/07/2023
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	01/07/2023	01/07/2023
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	01/07/2023	01/07/2023
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	01/07/2023	01/07/2023
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	01/07/2023	01/07/2023
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	01/07/2023	01/07/2023
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	01/07/2023	01/07/2023
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	01/07/2023	01/07/2023
11731	2022	All other reportable incidents	47265	Dual-chamber implantable pacemaker, rate-responsive	CT141	Electrodes/Leads and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	29/11/2022	29/11/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	29/11/2022	29/11/2022
					CT145	Cardiovascular devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	29/11/2022	29/11/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	29/11/2022	29/11/2022
					CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	29/11/2022	29/11/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	29/11/2022	29/11/2022
					CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	29/11/2022	29/11/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	29/11/2022	29/11/2022
					CT233	Surgical	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	29/11/2022	29/11/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	29/11/2022	29/11/2022
					CT2406	Implantable	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	29/11/2022	29/11/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	29/11/2022	29/11/2022
					CT2407	Active implantable	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	29/11/2022	29/11/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	29/11/2022	29/11/2022
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	29/11/2022	29/11/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	29/11/2022	29/11/2022
					CT2727	Active implant devices/ancillaries	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	29/11/2022	29/11/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	29/11/2022	29/11/2022
					CT2750	Implantable cardiac electrical stimulators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	29/11/2022	29/11/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	29/11/2022	29/11/2022
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	29/11/2022	29/11/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	29/11/2022	29/11/2022
					CT2985	Active	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	29/11/2022	29/11/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	29/11/2022	29/11/2022
					CT3001	Implantable metal	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	29/11/2022	29/11/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	29/11/2022	29/11/2022
					CT3002	Implantable synthetic polymer	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	29/11/2022	29/11/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	29/11/2022	29/11/2022
					CT3006	Implantable inorganic material	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	29/11/2022	29/11/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	29/11/2022	29/11/2022
					CT314	Electrically powered	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	29/11/2022	29/11/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	29/11/2022	29/11/2022
					CT321	Long-term surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	29/11/2022	29/11/2022

							F1901	Additional Surgery	A24	No Apparent Device Problem	29/11/2022	29/11/2022
					CT449	Generators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	29/11/2022	29/11/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	29/11/2022	29/11/2022
					CT981	Single use	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	29/11/2022	29/11/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	29/11/2022	29/11/2022
					CT983	Surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	29/11/2022	29/11/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	29/11/2022	29/11/2022
11732	2022	All other reportable incidents	47265	Dual-chamber implantable pacemaker, rate-responsive	CT141	Electrodes/Leads and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	30/11/2022	30/11/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	30/11/2022	30/11/2022
					CT145	Cardiovascular devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	30/11/2022	30/11/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	30/11/2022	30/11/2022
					CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	30/11/2022	30/11/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	30/11/2022	30/11/2022
					CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	30/11/2022	30/11/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	30/11/2022	30/11/2022
					CT233	Surgical	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	30/11/2022	30/11/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	30/11/2022	30/11/2022
					CT2406	Implantable	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	30/11/2022	30/11/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	30/11/2022	30/11/2022
					CT2407	Active implantable	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	30/11/2022	30/11/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	30/11/2022	30/11/2022
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	30/11/2022	30/11/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	30/11/2022	30/11/2022
					CT2727	Active implant devices/ancillaries	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	30/11/2022	30/11/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	30/11/2022	30/11/2022
					CT2750	Implantable cardiac electrical stimulators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	30/11/2022	30/11/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	30/11/2022	30/11/2022
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	30/11/2022	30/11/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	30/11/2022	30/11/2022
					CT2985	Active	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	30/11/2022	30/11/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	30/11/2022	30/11/2022
					CT3001	Implantable metal	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	30/11/2022	30/11/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	30/11/2022	30/11/2022
					CT3002	Implantable synthetic polymer	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	30/11/2022	30/11/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	30/11/2022	30/11/2022
					CT3006	Implantable inorganic material	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	30/11/2022	30/11/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	30/11/2022	30/11/2022
					CT314	Electrically powered	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	30/11/2022	30/11/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	30/11/2022	30/11/2022
					CT321	Long-term surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	30/11/2022	30/11/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	30/11/2022	30/11/2022
					CT449	Generators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	30/11/2022	30/11/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	30/11/2022	30/11/2022
					CT981	Single use	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	30/11/2022	30/11/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	30/11/2022	30/11/2022
					CT983	Surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	30/11/2022	30/11/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	30/11/2022	30/11/2022
11733	2022		47265	Dual-chamber implantable pacemaker, rate-responsive	CT141	Electrodes/Leads and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	01/12/2022	01/12/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	01/12/2022	01/12/2022
					CT145	Cardiovascular devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	01/12/2022	01/12/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	01/12/2022	01/12/2022
					CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	01/12/2022	01/12/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	01/12/2022	01/12/2022
					CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	01/12/2022	01/12/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	01/12/2022	01/12/2022
					CT233	Surgical	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	01/12/2022	01/12/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	01/12/2022	01/12/2022
					CT2406	Implantable	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	01/12/2022	01/12/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	01/12/2022	01/12/2022
					CT2407	Active implantable	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	01/12/2022	01/12/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	01/12/2022	01/12/2022
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	01/12/2022	01/12/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	01/12/2022	01/12/2022
					CT2727	Active implant devices/ancillaries	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	01/12/2022	01/12/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	01/12/2022	01/12/2022
					CT2750	Implantable cardiac electrical stimulators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	01/12/2022	01/12/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	01/12/2022	01/12/2022
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	01/12/2022	01/12/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	01/12/2022	01/12/2022
					CT2985	Active	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	01/12/2022	01/12/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	01/12/2022	01/12/2022
					CT3001	Implantable metal	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	01/12/2022	01/12/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	01/12/2022	01/12/2022
					CT3002	Implantable synthetic polymer	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	01/12/2022	01/12/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	01/12/2022	01/12/2022
					CT3006	Implantable inorganic material	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	01/12/2022	01/12/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	01/12/2022	01/12/2022
					CT314	Electrically powered	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	01/12/2022	01/12/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	01/12/2022	01/12/2022
					CT321	Long-term surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	01/12/2022	01/12/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	01/12/2022	01/12/2022
					CT449	Generators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	01/12/2022	01/12/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	01/12/2022	01/12/2022
					CT981	Single use	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	01/12/2022	01/12/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	01/12/2022	01/12/2022
					CT983	Surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	01/12/2022	01/12/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	01/12/2022	01/12/2022
11734	2022	All other reportable incidents	35223	Endocardial/interventricular septal pacing lead	CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A0401	Break		

									A070909	Oversensing		
							F22	Unexpected Diagnostic Intervention	A0401	Break		
									A070909	Oversensing		
11735	2022		35223	Endocardial/interventricular septal pacing lead	CT1684	Cardiac pacemakers and associated devices	F26	No Health Consequences or Impact	A040101	Fracture		
									A07	Electrical /Electronic Property Problem		
									A072201	High Impedance		
11736	2022		47263	Cardiac resynchronization therapy implantable pacemaker	CT1684	Cardiac pacemakers and associated devices	E01	Nervous System	A160602	Premature End-of-Life Indicator	08/12/2022	
							F19	Surgical Intervention	A160602	Premature End-of-Life Indicator	08/12/2022	
							F26	No Health Consequences or Impact	A160602	Premature End-of-Life Indicator	08/12/2022	
					CT1685	Cardiac pacemakers	E01	Nervous System	A160602	Premature End-of-Life Indicator	08/12/2022	
							F19	Surgical Intervention	A160602	Premature End-of-Life Indicator	08/12/2022	
							F26	No Health Consequences or Impact	A160602	Premature End-of-Life Indicator	08/12/2022	
					CT257	Implantable cardiac pacemakers	E01	Nervous System	A160602	Premature End-of-Life Indicator	08/12/2022	
							F19	Surgical Intervention	A160602	Premature End-of-Life Indicator	08/12/2022	
							F26	No Health Consequences or Impact	A160602	Premature End-of-Life Indicator	08/12/2022	
					CT2866	Implantable cardiac electrical stimulators	E01	Nervous System	A160602	Premature End-of-Life Indicator	08/12/2022	
							F19	Surgical Intervention	A160602	Premature End-of-Life Indicator	08/12/2022	
							F26	No Health Consequences or Impact	A160602	Premature End-of-Life Indicator	08/12/2022	
11737	2022	All other reportable incidents	47265	Dual-chamber implantable pacemaker, rate-responsive	CT141	Electrodes/Leads and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A071102	Failure to Interrogate	01/03/2023	01/03/2023
									A071204	No Pacing	01/03/2023	01/03/2023
							F1901	Additional Surgery	A071102	Failure to Interrogate	01/03/2023	01/03/2023
									A071204	No Pacing	01/03/2023	01/03/2023
					CT145	Cardiovascular devices	E2403	No Clinical Signs, Symptoms or Conditions	A071102	Failure to Interrogate	01/03/2023	01/03/2023
									A071204	No Pacing	01/03/2023	01/03/2023
							F1901	Additional Surgery	A071102	Failure to Interrogate	01/03/2023	01/03/2023
									A071204	No Pacing	01/03/2023	01/03/2023
					CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A071102	Failure to Interrogate	01/03/2023	01/03/2023
									A071204	No Pacing	01/03/2023	01/03/2023
							F1901	Additional Surgery	A071102	Failure to Interrogate	01/03/2023	01/03/2023
									A071204	No Pacing	01/03/2023	01/03/2023
					CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A071102	Failure to Interrogate	01/03/2023	01/03/2023
									A071204	No Pacing	01/03/2023	01/03/2023
							F1901	Additional Surgery	A071102	Failure to Interrogate	01/03/2023	01/03/2023
									A071204	No Pacing	01/03/2023	01/03/2023
					CT233	Surgical	E2403	No Clinical Signs, Symptoms or Conditions	A071102	Failure to Interrogate	01/03/2023	01/03/2023
									A071204	No Pacing	01/03/2023	01/03/2023
							F1901	Additional Surgery	A071102	Failure to Interrogate	01/03/2023	01/03/2023
									A071204	No Pacing	01/03/2023	01/03/2023
					CT2406	Implantable	E2403	No Clinical Signs, Symptoms or Conditions	A071102	Failure to Interrogate	01/03/2023	01/03/2023
									A071204	No Pacing	01/03/2023	01/03/2023
							F1901	Additional Surgery	A071102	Failure to Interrogate	01/03/2023	01/03/2023
									A071204	No Pacing	01/03/2023	01/03/2023
					CT2407	Active implantable	E2403	No Clinical Signs, Symptoms or Conditions	A071102	Failure to Interrogate	01/03/2023	01/03/2023
									A071204	No Pacing	01/03/2023	01/03/2023
							F1901	Additional Surgery	A071102	Failure to Interrogate	01/03/2023	01/03/2023
									A071204	No Pacing	01/03/2023	01/03/2023
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A071102	Failure to Interrogate	01/03/2023	01/03/2023
									A071204	No Pacing	01/03/2023	01/03/2023
							F1901	Additional Surgery	A071102	Failure to Interrogate	01/03/2023	01/03/2023
									A071204	No Pacing	01/03/2023	01/03/2023
					CT2727	Active implant devices/ancillaries	E2403	No Clinical Signs, Symptoms or Conditions	A071102	Failure to Interrogate	01/03/2023	01/03/2023
									A071204	No Pacing	01/03/2023	01/03/2023
							F1901	Additional Surgery	A071102	Failure to Interrogate	01/03/2023	01/03/2023
									A071204	No Pacing	01/03/2023	01/03/2023
					CT2750	Implantable cardiac electrical stimulators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A071102	Failure to Interrogate	01/03/2023	01/03/2023
									A071204	No Pacing	01/03/2023	01/03/2023
							F1901	Additional Surgery	A071102	Failure to Interrogate	01/03/2023	01/03/2023
									A071204	No Pacing	01/03/2023	01/03/2023
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A071102	Failure to Interrogate	01/03/2023	01/03/2023
									A071204	No Pacing	01/03/2023	01/03/2023
							F1901	Additional Surgery	A071102	Failure to Interrogate	01/03/2023	01/03/2023
									A071204	No Pacing	01/03/2023	01/03/2023
					CT2985	Active	E2403	No Clinical Signs, Symptoms or Conditions	A071102	Failure to Interrogate	01/03/2023	01/03/2023
									A071204	No Pacing	01/03/2023	01/03/2023
							F1901	Additional Surgery	A071102	Failure to Interrogate	01/03/2023	01/03/2023
									A071204	No Pacing	01/03/2023	01/03/2023
					CT3001	Implantable metal	E2403	No Clinical Signs, Symptoms or Conditions	A071102	Failure to Interrogate	01/03/2023	01/03/2023
									A071204	No Pacing	01/03/2023	01/03/2023
							F1901	Additional Surgery	A071102	Failure to Interrogate	01/03/2023	01/03/2023
									A071204	No Pacing	01/03/2023	01/03/2023
					CT3002	Implantable synthetic polymer	E2403	No Clinical Signs, Symptoms or Conditions	A071102	Failure to Interrogate	01/03/2023	01/03/2023
									A071204	No Pacing	01/03/2023	01/03/2023
							F1901	Additional Surgery	A071102	Failure to Interrogate	01/03/2023	01/03/2023
									A071204	No Pacing	01/03/2023	01/03/2023
					CT3006	Implantable inorganic material	E2403	No Clinical Signs, Symptoms or Conditions	A071102	Failure to Interrogate	01/03/2023	01/03/2023
									A071204	No Pacing	01/03/2023	01/03/2023
							F1901	Additional Surgery	A071102	Failure to Interrogate	01/03/2023	01/03/2023
									A071204	No Pacing	01/03/2023	01/03/2023
					CT314	Electrically powered	E2403	No Clinical Signs, Symptoms or Conditions	A071102	Failure to Interrogate	01/03/2023	01/03/2023
									A071204	No Pacing	01/03/2023	01/03/2023
							F1901	Additional Surgery	A071102	Failure to Interrogate	01/03/2023	01/03/2023
									A071204	No Pacing	01/03/2023	01/03/2023
					CT321	Long-term surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A071102	Failure to Interrogate	01/03/2023	01/03/2023
									A071204	No Pacing	01/03/2023	01/03/2023
							F1901	Additional Surgery	A071102	Failure to Interrogate	01/03/2023	01/03/2023
									A071204	No Pacing	01/03/2023	01/03/2023
					CT449	Generators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A071102	Failure to Interrogate	01/03/2023	01/03/2023
									A071204	No Pacing	01/03/2023	01/03/2023

							F1901	Additional Surgery	A071102	Failure to Interrogate	01/03/2023	01/03/2023
									A071204	No Pacing	01/03/2023	01/03/2023
					CT981	Single use	E2403	No Clinical Signs, Symptoms or Conditions	A071102	Failure to Interrogate	01/03/2023	01/03/2023
									A071204	No Pacing	01/03/2023	01/03/2023
							F1901	Additional Surgery	A071102	Failure to Interrogate	01/03/2023	01/03/2023
									A071204	No Pacing	01/03/2023	01/03/2023
					CT983	Surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A071102	Failure to Interrogate	01/03/2023	01/03/2023
									A071204	No Pacing	01/03/2023	01/03/2023
							F1901	Additional Surgery	A071102	Failure to Interrogate	01/03/2023	01/03/2023
									A071204	No Pacing	01/03/2023	01/03/2023
11738	2022	All other reportable incidents	47265	Dual-chamber implantable pacemaker, rate-responsive	CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
							F24	Insufficient Health Impact Information	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
							F24	Insufficient Health Impact Information	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
							F24	Insufficient Health Impact Information	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
11739	2022	All other reportable incidents	47265	Dual-chamber implantable pacemaker, rate-responsive	CT141	Electrodes/Leads and associated devices	E0112	Dizziness	A071102	Failure to Interrogate	30/11/2022	30/11/2022
									A071204	No Pacing	30/11/2022	30/11/2022
							F1901	Additional Surgery	A071102	Failure to Interrogate	30/11/2022	30/11/2022
									A071204	No Pacing	30/11/2022	30/11/2022
					CT145	Cardiovascular devices	E0112	Dizziness	A071102	Failure to Interrogate	30/11/2022	30/11/2022
									A071204	No Pacing	30/11/2022	30/11/2022
							F1901	Additional Surgery	A071102	Failure to Interrogate	30/11/2022	30/11/2022
									A071204	No Pacing	30/11/2022	30/11/2022
					CT1684	Cardiac pacemakers and associated devices	E0112	Dizziness	A071102	Failure to Interrogate	30/11/2022	30/11/2022
									A071204	No Pacing	30/11/2022	30/11/2022
							F1901	Additional Surgery	A071102	Failure to Interrogate	30/11/2022	30/11/2022
									A071204	No Pacing	30/11/2022	30/11/2022
					CT1685	Cardiac pacemakers	E0112	Dizziness	A071102	Failure to Interrogate	30/11/2022	30/11/2022
									A071204	No Pacing	30/11/2022	30/11/2022
							F1901	Additional Surgery	A071102	Failure to Interrogate	30/11/2022	30/11/2022
									A071204	No Pacing	30/11/2022	30/11/2022
					CT233	Surgical	E0112	Dizziness	A071102	Failure to Interrogate	30/11/2022	30/11/2022
									A071204	No Pacing	30/11/2022	30/11/2022
							F1901	Additional Surgery	A071102	Failure to Interrogate	30/11/2022	30/11/2022
									A071204	No Pacing	30/11/2022	30/11/2022
					CT2406	Implantable	E0112	Dizziness	A071102	Failure to Interrogate	30/11/2022	30/11/2022
									A071204	No Pacing	30/11/2022	30/11/2022
							F1901	Additional Surgery	A071102	Failure to Interrogate	30/11/2022	30/11/2022
									A071204	No Pacing	30/11/2022	30/11/2022
					CT2407	Active implantable	E0112	Dizziness	A071102	Failure to Interrogate	30/11/2022	30/11/2022
									A071204	No Pacing	30/11/2022	30/11/2022
							F1901	Additional Surgery	A071102	Failure to Interrogate	30/11/2022	30/11/2022
									A071204	No Pacing	30/11/2022	30/11/2022
					CT257	Implantable cardiac pacemakers	E0112	Dizziness	A071102	Failure to Interrogate	30/11/2022	30/11/2022
									A071204	No Pacing	30/11/2022	30/11/2022
							F1901	Additional Surgery	A071102	Failure to Interrogate	30/11/2022	30/11/2022
									A071204	No Pacing	30/11/2022	30/11/2022
					CT2727	Active implant devices/ancillaries	E0112	Dizziness	A071102	Failure to Interrogate	30/11/2022	30/11/2022
									A071204	No Pacing	30/11/2022	30/11/2022
							F1901	Additional Surgery	A071102	Failure to Interrogate	30/11/2022	30/11/2022
									A071204	No Pacing	30/11/2022	30/11/2022
					CT2750	Implantable cardiac electrical stimulators and associated devices	E0112	Dizziness	A071102	Failure to Interrogate	30/11/2022	30/11/2022
									A071204	No Pacing	30/11/2022	30/11/2022
							F1901	Additional Surgery	A071102	Failure to Interrogate	30/11/2022	30/11/2022
									A071204	No Pacing	30/11/2022	30/11/2022
					CT2866	Implantable cardiac electrical stimulators	E0112	Dizziness	A071102	Failure to Interrogate	30/11/2022	30/11/2022
									A071204	No Pacing	30/11/2022	30/11/2022
							F1901	Additional Surgery	A071102	Failure to Interrogate	30/11/2022	30/11/2022
									A071204	No Pacing	30/11/2022	30/11/2022
					CT2985	Active	E0112	Dizziness	A071102	Failure to Interrogate	30/11/2022	30/11/2022
									A071204	No Pacing	30/11/2022	30/11/2022
							F1901	Additional Surgery	A071102	Failure to Interrogate	30/11/2022	30/11/2022
									A071204	No Pacing	30/11/2022	30/11/2022
					CT3001	Implantable metal	E0112	Dizziness	A071102	Failure to Interrogate	30/11/2022	30/11/2022
									A071204	No Pacing	30/11/2022	30/11/2022
							F1901	Additional Surgery	A071102	Failure to Interrogate	30/11/2022	30/11/2022
									A071204	No Pacing	30/11/2022	30/11/2022
					CT3002	Implantable synthetic polymer	E0112	Dizziness	A071102	Failure to Interrogate	30/11/2022	30/11/2022
									A071204	No Pacing	30/11/2022	30/11/2022
							F1901	Additional Surgery	A071102	Failure to Interrogate	30/11/2022	30/11/2022
									A071204	No Pacing	30/11/2022	30/11/2022
					CT3006	Implantable inorganic material	E0112	Dizziness	A071102	Failure to Interrogate	30/11/2022	30/11/2022
									A071204	No Pacing	30/11/2022	30/11/2022
							F1901	Additional Surgery	A071102	Failure to Interrogate	30/11/2022	30/11/2022
									A071204	No Pacing	30/11/2022	30/11/2022
					CT314	Electrically powered	E0112	Dizziness	A071102	Failure to Interrogate	30/11/2022	30/11/2022
									A071204	No Pacing	30/11/2022	30/11/2022
							F1901	Additional Surgery	A071102	Failure to Interrogate	30/11/2022	30/11/2022
									A071204	No Pacing	30/11/2022	30/11/2022
					CT321	Long-term surgical invasive	E0112	Dizziness	A071102	Failure to Interrogate	30/11/2022	30/11/2022
									A071204	No Pacing	30/11/2022	30/11/2022

							F1901	Additional Surgery	A071102	Failure to Interrogate	30/11/2022	30/11/2022
									A071204	No Pacing	30/11/2022	30/11/2022
					CT449	Generators and associated devices	E0112	Dizziness	A071102	Failure to Interrogate	30/11/2022	30/11/2022
									A071204	No Pacing	30/11/2022	30/11/2022
							F1901	Additional Surgery	A071102	Failure to Interrogate	30/11/2022	30/11/2022
									A071204	No Pacing	30/11/2022	30/11/2022
					CT981	Single use	E0112	Dizziness	A071102	Failure to Interrogate	30/11/2022	30/11/2022
									A071204	No Pacing	30/11/2022	30/11/2022
							F1901	Additional Surgery	A071102	Failure to Interrogate	30/11/2022	30/11/2022
									A071204	No Pacing	30/11/2022	30/11/2022
					CT983	Surgical invasive	E0112	Dizziness	A071102	Failure to Interrogate	30/11/2022	30/11/2022
									A071204	No Pacing	30/11/2022	30/11/2022
							F1901	Additional Surgery	A071102	Failure to Interrogate	30/11/2022	30/11/2022
									A071204	No Pacing	30/11/2022	30/11/2022
11740	2022		47265	Dual-chamber implantable pacemaker, rate-responsive	CT141	Electrodes/Leads and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	27/09/2022	27/09/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	27/09/2022	27/09/2022
					CT145	Cardiovascular devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	27/09/2022	27/09/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	27/09/2022	27/09/2022
					CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	27/09/2022	27/09/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	27/09/2022	27/09/2022
					CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	27/09/2022	27/09/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	27/09/2022	27/09/2022
					CT233	Surgical	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	27/09/2022	27/09/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	27/09/2022	27/09/2022
					CT2406	Implantable	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	27/09/2022	27/09/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	27/09/2022	27/09/2022
					CT2407	Active implantable	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	27/09/2022	27/09/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	27/09/2022	27/09/2022
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	27/09/2022	27/09/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	27/09/2022	27/09/2022
					CT2727	Active implant devices/ancillaries	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	27/09/2022	27/09/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	27/09/2022	27/09/2022
					CT2750	Implantable cardiac electrical stimulators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	27/09/2022	27/09/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	27/09/2022	27/09/2022
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	27/09/2022	27/09/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	27/09/2022	27/09/2022
					CT2985	Active	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	27/09/2022	27/09/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	27/09/2022	27/09/2022
					CT3001	Implantable metal	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	27/09/2022	27/09/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	27/09/2022	27/09/2022
					CT3002	Implantable synthetic polymer	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	27/09/2022	27/09/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	27/09/2022	27/09/2022
					CT3006	Implantable inorganic material	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	27/09/2022	27/09/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	27/09/2022	27/09/2022
					CT314	Electrically powered	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	27/09/2022	27/09/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	27/09/2022	27/09/2022
					CT321	Long-term surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	27/09/2022	27/09/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	27/09/2022	27/09/2022
					CT449	Generators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	27/09/2022	27/09/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	27/09/2022	27/09/2022
					CT981	Single use	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	27/09/2022	27/09/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	27/09/2022	27/09/2022
					CT983	Surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	27/09/2022	27/09/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	27/09/2022	27/09/2022
11741	2022	All other reportable incidents	47265	Dual-chamber implantable pacemaker, rate-responsive	CT141	Electrodes/Leads and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	29/11/2022	29/11/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	29/11/2022	29/11/2022
					CT145	Cardiovascular devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	29/11/2022	29/11/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	29/11/2022	29/11/2022
					CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	29/11/2022	29/11/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	29/11/2022	29/11/2022
					CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	29/11/2022	29/11/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	29/11/2022	29/11/2022
					CT233	Surgical	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	29/11/2022	29/11/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	29/11/2022	29/11/2022
					CT2406	Implantable	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	29/11/2022	29/11/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	29/11/2022	29/11/2022
					CT2407	Active implantable	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	29/11/2022	29/11/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	29/11/2022	29/11/2022
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	29/11/2022	29/11/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	29/11/2022	29/11/2022
					CT2727	Active implant devices/ancillaries	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	29/11/2022	29/11/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	29/11/2022	29/11/2022
					CT2750	Implantable cardiac electrical stimulators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	29/11/2022	29/11/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	29/11/2022	29/11/2022
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	29/11/2022	29/11/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	29/11/2022	29/11/2022
					CT2985	Active	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	29/11/2022	29/11/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	29/11/2022	29/11/2022
					CT3001	Implantable metal	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	29/11/2022	29/11/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	29/11/2022	29/11/2022
					CT3002	Implantable synthetic polymer	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	29/11/2022	29/11/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	29/11/2022	29/11/2022
					CT3006	Implantable inorganic material	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	29/11/2022	29/11/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	29/11/2022	29/11/2022
					CT314	Electrically powered	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	29/11/2022	29/11/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	29/11/2022	29/11/2022
					CT321	Long-term surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	29/11/2022	29/11/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	29/11/2022	29/11/2022

					CT449	Generators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	29/11/2022	29/11/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	29/11/2022	29/11/2022
					CT981	Single use	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	29/11/2022	29/11/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	29/11/2022	29/11/2022
					CT983	Surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	29/11/2022	29/11/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	29/11/2022	29/11/2022
11742	2022	All other reportable incidents	47265	Dual-chamber implantable pacemaker, rate-responsive	CT141	Electrodes/Leads and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	26/10/2022	26/10/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	26/10/2022	26/10/2022
					CT145	Cardiovascular devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	26/10/2022	26/10/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	26/10/2022	26/10/2022
					CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	26/10/2022	26/10/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	26/10/2022	26/10/2022
					CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	26/10/2022	26/10/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	26/10/2022	26/10/2022
					CT233	Surgical	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	26/10/2022	26/10/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	26/10/2022	26/10/2022
					CT2406	Implantable	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	26/10/2022	26/10/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	26/10/2022	26/10/2022
					CT2407	Active implantable	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	26/10/2022	26/10/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	26/10/2022	26/10/2022
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	26/10/2022	26/10/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	26/10/2022	26/10/2022
					CT2727	Active implant devices/ancillaries	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	26/10/2022	26/10/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	26/10/2022	26/10/2022
					CT2750	Implantable cardiac electrical stimulators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	26/10/2022	26/10/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	26/10/2022	26/10/2022
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	26/10/2022	26/10/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	26/10/2022	26/10/2022
					CT2985	Active	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	26/10/2022	26/10/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	26/10/2022	26/10/2022
					CT3001	Implantable metal	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	26/10/2022	26/10/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	26/10/2022	26/10/2022
					CT3002	Implantable synthetic polymer	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	26/10/2022	26/10/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	26/10/2022	26/10/2022
					CT3006	Implantable inorganic material	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	26/10/2022	26/10/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	26/10/2022	26/10/2022
					CT314	Electrically powered	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	26/10/2022	26/10/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	26/10/2022	26/10/2022
					CT321	Long-term surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	26/10/2022	26/10/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	26/10/2022	26/10/2022
					CT449	Generators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	26/10/2022	26/10/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	26/10/2022	26/10/2022
					CT981	Single use	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	26/10/2022	26/10/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	26/10/2022	26/10/2022
					CT983	Surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	26/10/2022	26/10/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	26/10/2022	26/10/2022
11743	2022	All other reportable incidents	47265	Dual-chamber implantable pacemaker, rate-responsive	CT141	Electrodes/Leads and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	09/11/2022	09/11/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	09/11/2022	09/11/2022
					CT145	Cardiovascular devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	09/11/2022	09/11/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	09/11/2022	09/11/2022
					CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	09/11/2022	09/11/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	09/11/2022	09/11/2022
					CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	09/11/2022	09/11/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	09/11/2022	09/11/2022
					CT233	Surgical	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	09/11/2022	09/11/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	09/11/2022	09/11/2022
					CT2406	Implantable	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	09/11/2022	09/11/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	09/11/2022	09/11/2022
					CT2407	Active implantable	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	09/11/2022	09/11/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	09/11/2022	09/11/2022
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	09/11/2022	09/11/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	09/11/2022	09/11/2022
					CT2727	Active implant devices/ancillaries	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	09/11/2022	09/11/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	09/11/2022	09/11/2022
					CT2750	Implantable cardiac electrical stimulators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	09/11/2022	09/11/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	09/11/2022	09/11/2022
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	09/11/2022	09/11/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	09/11/2022	09/11/2022
					CT2985	Active	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	09/11/2022	09/11/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	09/11/2022	09/11/2022
					CT3001	Implantable metal	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	09/11/2022	09/11/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	09/11/2022	09/11/2022
					CT3002	Implantable synthetic polymer	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	09/11/2022	09/11/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	09/11/2022	09/11/2022
					CT3006	Implantable inorganic material	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	09/11/2022	09/11/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	09/11/2022	09/11/2022
					CT314	Electrically powered	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	09/11/2022	09/11/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	09/11/2022	09/11/2022
					CT321	Long-term surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	09/11/2022	09/11/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	09/11/2022	09/11/2022
					CT449	Generators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	09/11/2022	09/11/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	09/11/2022	09/11/2022
					CT981	Single use	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	09/11/2022	09/11/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	09/11/2022	09/11/2022
					CT983	Surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	09/11/2022	09/11/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	09/11/2022	09/11/2022
11744	2022	All other reportable incidents	47265	Dual-chamber implantable pacemaker, rate-responsive	CT141	Electrodes/Leads and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	09/11/2022	09/11/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	09/11/2022	09/11/2022

					CT145	Cardiovascular devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	09/11/2022	09/11/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	09/11/2022	09/11/2022
					CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	09/11/2022	09/11/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	09/11/2022	09/11/2022
					CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	09/11/2022	09/11/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	09/11/2022	09/11/2022
					CT233	Surgical	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	09/11/2022	09/11/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	09/11/2022	09/11/2022
					CT2406	Implantable	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	09/11/2022	09/11/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	09/11/2022	09/11/2022
					CT2407	Active implantable	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	09/11/2022	09/11/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	09/11/2022	09/11/2022
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	09/11/2022	09/11/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	09/11/2022	09/11/2022
					CT2727	Active implant devices/ancillaries	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	09/11/2022	09/11/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	09/11/2022	09/11/2022
					CT2750	Implantable cardiac electrical stimulators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	09/11/2022	09/11/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	09/11/2022	09/11/2022
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	09/11/2022	09/11/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	09/11/2022	09/11/2022
					CT2985	Active	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	09/11/2022	09/11/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	09/11/2022	09/11/2022
					CT3001	Implantable metal	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	09/11/2022	09/11/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	09/11/2022	09/11/2022
					CT3002	Implantable synthetic polymer	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	09/11/2022	09/11/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	09/11/2022	09/11/2022
					CT3006	Implantable inorganic material	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	09/11/2022	09/11/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	09/11/2022	09/11/2022
					CT314	Electrically powered	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	09/11/2022	09/11/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	09/11/2022	09/11/2022
					CT321	Long-term surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	09/11/2022	09/11/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	09/11/2022	09/11/2022
					CT449	Generators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	09/11/2022	09/11/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	09/11/2022	09/11/2022
					CT981	Single use	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	09/11/2022	09/11/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	09/11/2022	09/11/2022
					CT983	Surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	09/11/2022	09/11/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	09/11/2022	09/11/2022
11745	2022	All other reportable incidents	47265	Dual-chamber implantable pacemaker, rate-responsive	CT141	Electrodes/Leads and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	16/11/2022	16/11/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	16/11/2022	16/11/2022
					CT145	Cardiovascular devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	16/11/2022	16/11/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	16/11/2022	16/11/2022
					CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	16/11/2022	16/11/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	16/11/2022	16/11/2022
					CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	16/11/2022	16/11/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	16/11/2022	16/11/2022
					CT233	Surgical	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	16/11/2022	16/11/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	16/11/2022	16/11/2022
					CT2406	Implantable	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	16/11/2022	16/11/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	16/11/2022	16/11/2022
					CT2407	Active implantable	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	16/11/2022	16/11/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	16/11/2022	16/11/2022
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	16/11/2022	16/11/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	16/11/2022	16/11/2022
					CT2727	Active implant devices/ancillaries	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	16/11/2022	16/11/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	16/11/2022	16/11/2022
					CT2750	Implantable cardiac electrical stimulators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	16/11/2022	16/11/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	16/11/2022	16/11/2022
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	16/11/2022	16/11/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	16/11/2022	16/11/2022
					CT2985	Active	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	16/11/2022	16/11/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	16/11/2022	16/11/2022
					CT3001	Implantable metal	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	16/11/2022	16/11/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	16/11/2022	16/11/2022
					CT3002	Implantable synthetic polymer	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	16/11/2022	16/11/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	16/11/2022	16/11/2022
					CT3006	Implantable inorganic material	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	16/11/2022	16/11/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	16/11/2022	16/11/2022
					CT314	Electrically powered	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	16/11/2022	16/11/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	16/11/2022	16/11/2022
					CT321	Long-term surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	16/11/2022	16/11/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	16/11/2022	16/11/2022
					CT449	Generators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	16/11/2022	16/11/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	16/11/2022	16/11/2022
					CT981	Single use	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	16/11/2022	16/11/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	16/11/2022	16/11/2022
					CT983	Surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	16/11/2022	16/11/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	16/11/2022	16/11/2022
11746	2022	All other reportable incidents	47265	Dual-chamber implantable pacemaker, rate-responsive	CT141	Electrodes/Leads and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	28/11/2022	28/11/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	28/11/2022	28/11/2022
					CT145	Cardiovascular devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	28/11/2022	28/11/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	28/11/2022	28/11/2022
					CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	28/11/2022	28/11/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	28/11/2022	28/11/2022
					CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	28/11/2022	28/11/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	28/11/2022	28/11/2022
					CT233	Surgical	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	28/11/2022	28/11/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	28/11/2022	28/11/2022

					CT2406	Implantable	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	28/11/2022	28/11/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	28/11/2022	28/11/2022
					CT2407	Active implantable	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	28/11/2022	28/11/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	28/11/2022	28/11/2022
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	28/11/2022	28/11/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	28/11/2022	28/11/2022
					CT2727	Active implant devices/ancillaries	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	28/11/2022	28/11/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	28/11/2022	28/11/2022
					CT2750	Implantable cardiac electrical stimulators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	28/11/2022	28/11/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	28/11/2022	28/11/2022
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	28/11/2022	28/11/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	28/11/2022	28/11/2022
					CT2985	Active	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	28/11/2022	28/11/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	28/11/2022	28/11/2022
					CT3001	Implantable metal	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	28/11/2022	28/11/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	28/11/2022	28/11/2022
					CT3002	Implantable synthetic polymer	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	28/11/2022	28/11/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	28/11/2022	28/11/2022
					CT3006	Implantable inorganic material	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	28/11/2022	28/11/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	28/11/2022	28/11/2022
					CT314	Electrically powered	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	28/11/2022	28/11/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	28/11/2022	28/11/2022
					CT321	Long-term surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	28/11/2022	28/11/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	28/11/2022	28/11/2022
					CT449	Generators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	28/11/2022	28/11/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	28/11/2022	28/11/2022
					CT981	Single use	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	28/11/2022	28/11/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	28/11/2022	28/11/2022
					CT983	Surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	28/11/2022	28/11/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	28/11/2022	28/11/2022
11747	2022	All other reportable incidents	37265	Dual-chamber implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E2311	Discomfort	A070909	Oversensing		
									A071302	Inappropriate/Inadequate Shock/Stimulation		
							F23	Unexpected Medical Intervention	A070909	Oversensing		
									A071302	Inappropriate/Inadequate Shock/Stimulation		
11748	2022	All other reportable incidents	35852	Single-chamber implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	19/12/2022	19/12/2022
							F1901	Additional Surgery	A070504	Premature Discharge of Battery	19/12/2022	19/12/2022
11749	2022	All other reportable incidents	35223	Endocardial/interventricular septal pacing lead	CT1684	Cardiac pacemakers and associated devices	E0601	Arrhythmia	A070101	Failure to Capture		
							F26	No Health Consequences or Impact	A070101	Failure to Capture		
11750	2022	All other reportable incidents	47265	Dual-chamber implantable pacemaker, rate-responsive	CT141	Electrodes/Leads and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	31/10/2022	31/10/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	31/10/2022	31/10/2022
					CT145	Cardiovascular devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	31/10/2022	31/10/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	31/10/2022	31/10/2022
					CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	31/10/2022	31/10/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	31/10/2022	31/10/2022
					CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	31/10/2022	31/10/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	31/10/2022	31/10/2022
					CT233	Surgical	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	31/10/2022	31/10/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	31/10/2022	31/10/2022
					CT2406	Implantable	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	31/10/2022	31/10/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	31/10/2022	31/10/2022
					CT2407	Active implantable	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	31/10/2022	31/10/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	31/10/2022	31/10/2022
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	31/10/2022	31/10/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	31/10/2022	31/10/2022
					CT2727	Active implant devices/ancillaries	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	31/10/2022	31/10/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	31/10/2022	31/10/2022
					CT2750	Implantable cardiac electrical stimulators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	31/10/2022	31/10/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	31/10/2022	31/10/2022
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	31/10/2022	31/10/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	31/10/2022	31/10/2022
					CT2985	Active	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	31/10/2022	31/10/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	31/10/2022	31/10/2022
					CT3001	Implantable metal	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	31/10/2022	31/10/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	31/10/2022	31/10/2022
					CT3002	Implantable synthetic polymer	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	31/10/2022	31/10/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	31/10/2022	31/10/2022
					CT3006	Implantable inorganic material	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	31/10/2022	31/10/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	31/10/2022	31/10/2022
					CT314	Electrically powered	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	31/10/2022	31/10/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	31/10/2022	31/10/2022
					CT321	Long-term surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	31/10/2022	31/10/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	31/10/2022	31/10/2022
					CT449	Generators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	31/10/2022	31/10/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	31/10/2022	31/10/2022
					CT981	Single use	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	31/10/2022	31/10/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	31/10/2022	31/10/2022
					CT983	Surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	31/10/2022	31/10/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	31/10/2022	31/10/2022
11751	2022	All other reportable incidents	47265	Dual-chamber implantable pacemaker, rate-responsive	CT141	Electrodes/Leads and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	10/10/2022	10/10/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	10/10/2022	10/10/2022
					CT145	Cardiovascular devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	10/10/2022	10/10/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	10/10/2022	10/10/2022
					CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	10/10/2022	10/10/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	10/10/2022	10/10/2022
					CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	10/10/2022	10/10/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	10/10/2022	10/10/2022
					CT233	Surgical	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	10/10/2022	10/10/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	10/10/2022	10/10/2022

					CT2406	Implantable	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	10/10/2022	10/10/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	10/10/2022	10/10/2022
					CT2407	Active implantable	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	10/10/2022	10/10/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	10/10/2022	10/10/2022
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	10/10/2022	10/10/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	10/10/2022	10/10/2022
					CT2727	Active implant devices/ancillaries	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	10/10/2022	10/10/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	10/10/2022	10/10/2022
					CT2750	Implantable cardiac electrical stimulators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	10/10/2022	10/10/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	10/10/2022	10/10/2022
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	10/10/2022	10/10/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	10/10/2022	10/10/2022
					CT2985	Active	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	10/10/2022	10/10/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	10/10/2022	10/10/2022
					CT3001	Implantable metal	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	10/10/2022	10/10/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	10/10/2022	10/10/2022
					CT3002	Implantable synthetic polymer	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	10/10/2022	10/10/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	10/10/2022	10/10/2022
					CT3006	Implantable inorganic material	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	10/10/2022	10/10/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	10/10/2022	10/10/2022
					CT314	Electrically powered	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	10/10/2022	10/10/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	10/10/2022	10/10/2022
					CT321	Long-term surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	10/10/2022	10/10/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	10/10/2022	10/10/2022
					CT449	Generators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	10/10/2022	10/10/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	10/10/2022	10/10/2022
					CT981	Single use	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	10/10/2022	10/10/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	10/10/2022	10/10/2022
					CT983	Surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	10/10/2022	10/10/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	10/10/2022	10/10/2022
11752	2022	All other reportable incidents	47265	Dual-chamber implantable pacemaker, rate-responsive	CT141	Electrodes/Leads and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	12/11/2022	12/11/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	12/11/2022	12/11/2022
					CT145	Cardiovascular devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	12/11/2022	12/11/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	12/11/2022	12/11/2022
					CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	12/11/2022	12/11/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	12/11/2022	12/11/2022
					CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	12/11/2022	12/11/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	12/11/2022	12/11/2022
					CT233	Surgical	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	12/11/2022	12/11/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	12/11/2022	12/11/2022
					CT2406	Implantable	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	12/11/2022	12/11/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	12/11/2022	12/11/2022
					CT2407	Active implantable	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	12/11/2022	12/11/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	12/11/2022	12/11/2022
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	12/11/2022	12/11/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	12/11/2022	12/11/2022
					CT2727	Active implant devices/ancillaries	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	12/11/2022	12/11/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	12/11/2022	12/11/2022
					CT2750	Implantable cardiac electrical stimulators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	12/11/2022	12/11/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	12/11/2022	12/11/2022
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	12/11/2022	12/11/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	12/11/2022	12/11/2022
					CT2985	Active	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	12/11/2022	12/11/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	12/11/2022	12/11/2022
					CT3001	Implantable metal	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	12/11/2022	12/11/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	12/11/2022	12/11/2022
					CT3002	Implantable synthetic polymer	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	12/11/2022	12/11/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	12/11/2022	12/11/2022
					CT3006	Implantable inorganic material	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	12/11/2022	12/11/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	12/11/2022	12/11/2022
					CT314	Electrically powered	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	12/11/2022	12/11/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	12/11/2022	12/11/2022
					CT321	Long-term surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	12/11/2022	12/11/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	12/11/2022	12/11/2022
					CT449	Generators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	12/11/2022	12/11/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	12/11/2022	12/11/2022
					CT981	Single use	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	12/11/2022	12/11/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	12/11/2022	12/11/2022
					CT983	Surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	12/11/2022	12/11/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	12/11/2022	12/11/2022
11753	2022	All other reportable incidents	47265	Dual-chamber implantable pacemaker, rate-responsive	CT141	Electrodes/Leads and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	22/08/2022	22/08/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	22/08/2022	22/08/2022
					CT145	Cardiovascular devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	22/08/2022	22/08/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	22/08/2022	22/08/2022
					CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	22/08/2022	22/08/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	22/08/2022	22/08/2022
					CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	22/08/2022	22/08/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	22/08/2022	22/08/2022
					CT233	Surgical	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	22/08/2022	22/08/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	22/08/2022	22/08/2022
					CT2406	Implantable	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	22/08/2022	22/08/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	22/08/2022	22/08/2022
					CT2407	Active implantable	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	22/08/2022	22/08/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	22/08/2022	22/08/2022
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	22/08/2022	22/08/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	22/08/2022	22/08/2022
					CT2727	Active implant devices/ancillaries	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	22/08/2022	22/08/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	22/08/2022	22/08/2022

					CT2750	Implantable cardiac electrical stimulators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	22/08/2022	22/08/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	22/08/2022	22/08/2022
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	22/08/2022	22/08/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	22/08/2022	22/08/2022
					CT2985	Active	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	22/08/2022	22/08/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	22/08/2022	22/08/2022
					CT3001	Implantable metal	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	22/08/2022	22/08/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	22/08/2022	22/08/2022
					CT3002	Implantable synthetic polymer	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	22/08/2022	22/08/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	22/08/2022	22/08/2022
					CT3006	Implantable inorganic material	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	22/08/2022	22/08/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	22/08/2022	22/08/2022
					CT314	Electrically powered	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	22/08/2022	22/08/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	22/08/2022	22/08/2022
					CT321	Long-term surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	22/08/2022	22/08/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	22/08/2022	22/08/2022
					CT449	Generators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	22/08/2022	22/08/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	22/08/2022	22/08/2022
					CT981	Single use	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	22/08/2022	22/08/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	22/08/2022	22/08/2022
					CT983	Surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	22/08/2022	22/08/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	22/08/2022	22/08/2022
11754	2022	All other reportable incidents	47265	Dual-chamber implantable pacemaker, rate-responsive	CT141	Electrodes/Leads and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	05/09/2022	05/09/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	05/09/2022	05/09/2022
					CT145	Cardiovascular devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	05/09/2022	05/09/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	05/09/2022	05/09/2022
					CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	05/09/2022	05/09/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	05/09/2022	05/09/2022
					CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	05/09/2022	05/09/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	05/09/2022	05/09/2022
					CT233	Surgical	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	05/09/2022	05/09/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	05/09/2022	05/09/2022
					CT2406	Implantable	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	05/09/2022	05/09/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	05/09/2022	05/09/2022
					CT2407	Active implantable	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	05/09/2022	05/09/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	05/09/2022	05/09/2022
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	05/09/2022	05/09/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	05/09/2022	05/09/2022
					CT2727	Active implant devices/ancillaries	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	05/09/2022	05/09/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	05/09/2022	05/09/2022
					CT2750	Implantable cardiac electrical stimulators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	05/09/2022	05/09/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	05/09/2022	05/09/2022
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	05/09/2022	05/09/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	05/09/2022	05/09/2022
					CT2985	Active	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	05/09/2022	05/09/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	05/09/2022	05/09/2022
					CT3001	Implantable metal	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	05/09/2022	05/09/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	05/09/2022	05/09/2022
					CT3002	Implantable synthetic polymer	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	05/09/2022	05/09/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	05/09/2022	05/09/2022
					CT3006	Implantable inorganic material	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	05/09/2022	05/09/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	05/09/2022	05/09/2022
					CT314	Electrically powered	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	05/09/2022	05/09/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	05/09/2022	05/09/2022
					CT321	Long-term surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	05/09/2022	05/09/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	05/09/2022	05/09/2022
					CT449	Generators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	05/09/2022	05/09/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	05/09/2022	05/09/2022
					CT981	Single use	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	05/09/2022	05/09/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	05/09/2022	05/09/2022
					CT983	Surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	05/09/2022	05/09/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	05/09/2022	05/09/2022
11755	2022		47265	Dual-chamber implantable pacemaker, rate-responsive	CT141	Electrodes/Leads and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem		
							F1901	Additional Surgery	A24	No Apparent Device Problem		
					CT145	Cardiovascular devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem		
							F1901	Additional Surgery	A24	No Apparent Device Problem		
					CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem		
							F1901	Additional Surgery	A24	No Apparent Device Problem		
					CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem		
							F1901	Additional Surgery	A24	No Apparent Device Problem		
					CT233	Surgical	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem		
							F1901	Additional Surgery	A24	No Apparent Device Problem		
					CT2406	Implantable	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem		
							F1901	Additional Surgery	A24	No Apparent Device Problem		
					CT2407	Active implantable	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem		
							F1901	Additional Surgery	A24	No Apparent Device Problem		
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem		
							F1901	Additional Surgery	A24	No Apparent Device Problem		
					CT2727	Active implant devices/ancillaries	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem		
							F1901	Additional Surgery	A24	No Apparent Device Problem		
					CT2750	Implantable cardiac electrical stimulators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem		
							F1901	Additional Surgery	A24	No Apparent Device Problem		
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem		
							F1901	Additional Surgery	A24	No Apparent Device Problem		
					CT2985	Active	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem		
							F1901	Additional Surgery	A24	No Apparent Device Problem		
					CT3001	Implantable metal	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem		
							F1901	Additional Surgery	A24	No Apparent Device Problem		

					CT3002	Implantable synthetic polymer	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem		
							F1901	Additional Surgery	A24	No Apparent Device Problem		
					CT3006	Implantable inorganic material	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem		
							F1901	Additional Surgery	A24	No Apparent Device Problem		
					CT314	Electrically powered	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem		
							F1901	Additional Surgery	A24	No Apparent Device Problem		
					CT321	Long-term surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem		
							F1901	Additional Surgery	A24	No Apparent Device Problem		
					CT449	Generators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem		
							F1901	Additional Surgery	A24	No Apparent Device Problem		
					CT981	Single use	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem		
							F1901	Additional Surgery	A24	No Apparent Device Problem		
					CT983	Surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem		
							F1901	Additional Surgery	A24	No Apparent Device Problem		
11756	2022		47265	Dual-chamber implantable pacemaker, rate-responsive	CT141	Electrodes/Leads and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem		
							F1901	Additional Surgery	A24	No Apparent Device Problem		
					CT145	Cardiovascular devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem		
							F1901	Additional Surgery	A24	No Apparent Device Problem		
					CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem		
							F1901	Additional Surgery	A24	No Apparent Device Problem		
					CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem		
							F1901	Additional Surgery	A24	No Apparent Device Problem		
					CT233	Surgical	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem		
							F1901	Additional Surgery	A24	No Apparent Device Problem		
					CT2406	Implantable	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem		
							F1901	Additional Surgery	A24	No Apparent Device Problem		
					CT2407	Active implantable	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem		
							F1901	Additional Surgery	A24	No Apparent Device Problem		
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem		
							F1901	Additional Surgery	A24	No Apparent Device Problem		
					CT2727	Active implant devices/ancillaries	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem		
							F1901	Additional Surgery	A24	No Apparent Device Problem		
					CT2750	Implantable cardiac electrical stimulators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem		
							F1901	Additional Surgery	A24	No Apparent Device Problem		
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem		
							F1901	Additional Surgery	A24	No Apparent Device Problem		
					CT2985	Active	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem		
							F1901	Additional Surgery	A24	No Apparent Device Problem		
					CT3001	Implantable metal	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem		
							F1901	Additional Surgery	A24	No Apparent Device Problem		
					CT3002	Implantable synthetic polymer	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem		
							F1901	Additional Surgery	A24	No Apparent Device Problem		
					CT3006	Implantable inorganic material	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem		
							F1901	Additional Surgery	A24	No Apparent Device Problem		
					CT314	Electrically powered	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem		
							F1901	Additional Surgery	A24	No Apparent Device Problem		
					CT321	Long-term surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem		
							F1901	Additional Surgery	A24	No Apparent Device Problem		
					CT449	Generators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem		
							F1901	Additional Surgery	A24	No Apparent Device Problem		
					CT981	Single use	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem		
							F1901	Additional Surgery	A24	No Apparent Device Problem		
					CT983	Surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem		
							F1901	Additional Surgery	A24	No Apparent Device Problem		
11757	2022		47265	Dual-chamber implantable pacemaker, rate-responsive	CT141	Electrodes/Leads and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	20/11/2022	20/11/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	20/11/2022	20/11/2022
					CT145	Cardiovascular devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	20/11/2022	20/11/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	20/11/2022	20/11/2022
					CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	20/11/2022	20/11/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	20/11/2022	20/11/2022
					CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	20/11/2022	20/11/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	20/11/2022	20/11/2022
					CT233	Surgical	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	20/11/2022	20/11/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	20/11/2022	20/11/2022
					CT2406	Implantable	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	20/11/2022	20/11/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	20/11/2022	20/11/2022
					CT2407	Active implantable	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	20/11/2022	20/11/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	20/11/2022	20/11/2022
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	20/11/2022	20/11/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	20/11/2022	20/11/2022
					CT2727	Active implant devices/ancillaries	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	20/11/2022	20/11/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	20/11/2022	20/11/2022
					CT2750	Implantable cardiac electrical stimulators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	20/11/2022	20/11/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	20/11/2022	20/11/2022
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	20/11/2022	20/11/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	20/11/2022	20/11/2022
					CT2985	Active	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	20/11/2022	20/11/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	20/11/2022	20/11/2022
					CT3001	Implantable metal	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	20/11/2022	20/11/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	20/11/2022	20/11/2022
					CT3002	Implantable synthetic polymer	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	20/11/2022	20/11/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	20/11/2022	20/11/2022
					CT3006	Implantable inorganic material	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	20/11/2022	20/11/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	20/11/2022	20/11/2022
					CT314	Electrically powered	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	20/11/2022	20/11/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	20/11/2022	20/11/2022
					CT321	Long-term surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	20/11/2022	20/11/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	20/11/2022	20/11/2022

					CT449	Generators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	20/11/2022	20/11/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	20/11/2022	20/11/2022
					CT981	Single use	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	20/11/2022	20/11/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	20/11/2022	20/11/2022
					CT983	Surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	20/11/2022	20/11/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	20/11/2022	20/11/2022
11758	2022	All other reportable incidents	47265	Dual-chamber implantable pacemaker, rate-responsive	CT141	Electrodes/Leads and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	28/11/2022	28/11/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	28/11/2022	28/11/2022
					CT145	Cardiovascular devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	28/11/2022	28/11/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	28/11/2022	28/11/2022
					CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	28/11/2022	28/11/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	28/11/2022	28/11/2022
					CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	28/11/2022	28/11/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	28/11/2022	28/11/2022
					CT233	Surgical	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	28/11/2022	28/11/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	28/11/2022	28/11/2022
					CT2406	Implantable	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	28/11/2022	28/11/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	28/11/2022	28/11/2022
					CT2407	Active implantable	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	28/11/2022	28/11/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	28/11/2022	28/11/2022
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	28/11/2022	28/11/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	28/11/2022	28/11/2022
					CT2727	Active implant devices/ancillaries	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	28/11/2022	28/11/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	28/11/2022	28/11/2022
					CT2750	Implantable cardiac electrical stimulators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	28/11/2022	28/11/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	28/11/2022	28/11/2022
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	28/11/2022	28/11/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	28/11/2022	28/11/2022
					CT2985	Active	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	28/11/2022	28/11/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	28/11/2022	28/11/2022
					CT3001	Implantable metal	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	28/11/2022	28/11/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	28/11/2022	28/11/2022
					CT3002	Implantable synthetic polymer	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	28/11/2022	28/11/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	28/11/2022	28/11/2022
					CT3006	Implantable inorganic material	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	28/11/2022	28/11/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	28/11/2022	28/11/2022
					CT314	Electrically powered	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	28/11/2022	28/11/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	28/11/2022	28/11/2022
					CT321	Long-term surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	28/11/2022	28/11/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	28/11/2022	28/11/2022
					CT449	Generators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	28/11/2022	28/11/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	28/11/2022	28/11/2022
					CT981	Single use	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	28/11/2022	28/11/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	28/11/2022	28/11/2022
					CT983	Surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	28/11/2022	28/11/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	28/11/2022	28/11/2022
11759	2022		47265	Dual-chamber implantable pacemaker, rate-responsive	CT141	Electrodes/Leads and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	23/11/2022	23/11/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	23/11/2022	23/11/2022
					CT145	Cardiovascular devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	23/11/2022	23/11/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	23/11/2022	23/11/2022
					CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	23/11/2022	23/11/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	23/11/2022	23/11/2022
					CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	23/11/2022	23/11/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	23/11/2022	23/11/2022
					CT233	Surgical	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	23/11/2022	23/11/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	23/11/2022	23/11/2022
					CT2406	Implantable	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	23/11/2022	23/11/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	23/11/2022	23/11/2022
					CT2407	Active implantable	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	23/11/2022	23/11/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	23/11/2022	23/11/2022
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	23/11/2022	23/11/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	23/11/2022	23/11/2022
					CT2727	Active implant devices/ancillaries	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	23/11/2022	23/11/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	23/11/2022	23/11/2022
					CT2750	Implantable cardiac electrical stimulators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	23/11/2022	23/11/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	23/11/2022	23/11/2022
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	23/11/2022	23/11/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	23/11/2022	23/11/2022
					CT2985	Active	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	23/11/2022	23/11/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	23/11/2022	23/11/2022
					CT3001	Implantable metal	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	23/11/2022	23/11/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	23/11/2022	23/11/2022
					CT3002	Implantable synthetic polymer	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	23/11/2022	23/11/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	23/11/2022	23/11/2022
					CT3006	Implantable inorganic material	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	23/11/2022	23/11/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	23/11/2022	23/11/2022
					CT314	Electrically powered	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	23/11/2022	23/11/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	23/11/2022	23/11/2022
					CT321	Long-term surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	23/11/2022	23/11/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	23/11/2022	23/11/2022
					CT449	Generators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	23/11/2022	23/11/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	23/11/2022	23/11/2022
					CT981	Single use	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	23/11/2022	23/11/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	23/11/2022	23/11/2022
					CT983	Surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	23/11/2022	23/11/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	23/11/2022	23/11/2022
11760	2022	All other reportable incidents	47265	Dual-chamber implantable pacemaker, rate-responsive	CT141	Electrodes/Leads and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	26/09/2022	26/09/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	26/09/2022	26/09/2022

					CT145	Cardiovascular devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	26/09/2022	26/09/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	26/09/2022	26/09/2022
					CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	26/09/2022	26/09/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	26/09/2022	26/09/2022
					CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	26/09/2022	26/09/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	26/09/2022	26/09/2022
					CT233	Surgical	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	26/09/2022	26/09/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	26/09/2022	26/09/2022
					CT2406	Implantable	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	26/09/2022	26/09/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	26/09/2022	26/09/2022
					CT2407	Active implantable	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	26/09/2022	26/09/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	26/09/2022	26/09/2022
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	26/09/2022	26/09/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	26/09/2022	26/09/2022
					CT2727	Active implant devices/ancillaries	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	26/09/2022	26/09/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	26/09/2022	26/09/2022
					CT2750	Implantable cardiac electrical stimulators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	26/09/2022	26/09/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	26/09/2022	26/09/2022
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	26/09/2022	26/09/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	26/09/2022	26/09/2022
					CT2985	Active	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	26/09/2022	26/09/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	26/09/2022	26/09/2022
					CT3001	Implantable metal	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	26/09/2022	26/09/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	26/09/2022	26/09/2022
					CT3002	Implantable synthetic polymer	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	26/09/2022	26/09/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	26/09/2022	26/09/2022
					CT3006	Implantable inorganic material	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	26/09/2022	26/09/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	26/09/2022	26/09/2022
					CT314	Electrically powered	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	26/09/2022	26/09/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	26/09/2022	26/09/2022
					CT321	Long-term surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	26/09/2022	26/09/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	26/09/2022	26/09/2022
					CT449	Generators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	26/09/2022	26/09/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	26/09/2022	26/09/2022
					CT981	Single use	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	26/09/2022	26/09/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	26/09/2022	26/09/2022
					CT983	Surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	26/09/2022	26/09/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	26/09/2022	26/09/2022
11761	2022		47265	Dual-chamber implantable pacemaker, rate-responsive	CT141	Electrodes/Leads and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	20/10/2022	20/10/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	20/10/2022	20/10/2022
					CT145	Cardiovascular devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	20/10/2022	20/10/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	20/10/2022	20/10/2022
					CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	20/10/2022	20/10/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	20/10/2022	20/10/2022
					CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	20/10/2022	20/10/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	20/10/2022	20/10/2022
					CT233	Surgical	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	20/10/2022	20/10/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	20/10/2022	20/10/2022
					CT2406	Implantable	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	20/10/2022	20/10/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	20/10/2022	20/10/2022
					CT2407	Active implantable	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	20/10/2022	20/10/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	20/10/2022	20/10/2022
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	20/10/2022	20/10/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	20/10/2022	20/10/2022
					CT2727	Active implant devices/ancillaries	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	20/10/2022	20/10/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	20/10/2022	20/10/2022
					CT2750	Implantable cardiac electrical stimulators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	20/10/2022	20/10/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	20/10/2022	20/10/2022
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	20/10/2022	20/10/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	20/10/2022	20/10/2022
					CT2985	Active	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	20/10/2022	20/10/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	20/10/2022	20/10/2022
					CT3001	Implantable metal	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	20/10/2022	20/10/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	20/10/2022	20/10/2022
					CT3002	Implantable synthetic polymer	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	20/10/2022	20/10/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	20/10/2022	20/10/2022
					CT3006	Implantable inorganic material	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	20/10/2022	20/10/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	20/10/2022	20/10/2022
					CT314	Electrically powered	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	20/10/2022	20/10/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	20/10/2022	20/10/2022
					CT321	Long-term surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	20/10/2022	20/10/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	20/10/2022	20/10/2022
					CT449	Generators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	20/10/2022	20/10/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	20/10/2022	20/10/2022
					CT981	Single use	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	20/10/2022	20/10/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	20/10/2022	20/10/2022
					CT983	Surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	20/10/2022	20/10/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	20/10/2022	20/10/2022
11762	2022	All other reportable incidents	35852	Single-chamber implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F1901	Additional Surgery	A070504	Premature Discharge of Battery		
11763	2022	All other reportable incidents	47265	Dual-chamber implantable pacemaker, rate-responsive	CT141	Electrodes/Leads and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	17/10/2022	17/10/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	17/10/2022	17/10/2022
					CT145	Cardiovascular devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	17/10/2022	17/10/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	17/10/2022	17/10/2022
					CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	17/10/2022	17/10/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	17/10/2022	17/10/2022
					CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	17/10/2022	17/10/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	17/10/2022	17/10/2022

					CT233	Surgical	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	17/10/2022	17/10/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	17/10/2022	17/10/2022
					CT2406	Implantable	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	17/10/2022	17/10/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	17/10/2022	17/10/2022
					CT2407	Active implantable	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	17/10/2022	17/10/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	17/10/2022	17/10/2022
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	17/10/2022	17/10/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	17/10/2022	17/10/2022
					CT2727	Active implant devices/ancillaries	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	17/10/2022	17/10/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	17/10/2022	17/10/2022
					CT2750	Implantable cardiac electrical stimulators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	17/10/2022	17/10/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	17/10/2022	17/10/2022
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	17/10/2022	17/10/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	17/10/2022	17/10/2022
					CT2985	Active	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	17/10/2022	17/10/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	17/10/2022	17/10/2022
					CT3001	Implantable metal	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	17/10/2022	17/10/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	17/10/2022	17/10/2022
					CT3002	Implantable synthetic polymer	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	17/10/2022	17/10/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	17/10/2022	17/10/2022
					CT3006	Implantable inorganic material	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	17/10/2022	17/10/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	17/10/2022	17/10/2022
					CT314	Electrically powered	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	17/10/2022	17/10/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	17/10/2022	17/10/2022
					CT321	Long-term surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	17/10/2022	17/10/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	17/10/2022	17/10/2022
					CT449	Generators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	17/10/2022	17/10/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	17/10/2022	17/10/2022
					CT981	Single use	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	17/10/2022	17/10/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	17/10/2022	17/10/2022
					CT983	Surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	17/10/2022	17/10/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	17/10/2022	17/10/2022
11764	2022	Unanticipated serious deterioration in state of health	63188	Left-ventricular wireless pacing system sensor/transmitter	CT141	Electrodes/Leads and associated devices			A070504	Premature Discharge of Battery		
					CT145	Cardiovascular devices			A070504	Premature Discharge of Battery		
					CT1577	Wireless monitoring system transmitters			A070504	Premature Discharge of Battery		
					CT1684	Cardiac pacemakers and associated devices			A070504	Premature Discharge of Battery		
					CT2317	Transmitters/Receivers/Transceivers			A070504	Premature Discharge of Battery		
					CT233	Surgical			A070504	Premature Discharge of Battery		
					CT2406	Implantable			A070504	Premature Discharge of Battery		
					CT2407	Active implantable			A070504	Premature Discharge of Battery		
					CT255	Transducers and associated devices			A070504	Premature Discharge of Battery		
					CT2726	Physiological monitoring/measuring devices			A070504	Premature Discharge of Battery		
					CT2727	Active implant devices/ancillaries			A070504	Premature Discharge of Battery		
					CT2750	Implantable cardiac electrical stimulators and associated devices			A070504	Premature Discharge of Battery		
					CT2885	Physiological monitoring/measuring system transmitters/receivers/transceivers			A070504	Premature Discharge of Battery		
					CT2985	Active			A070504	Premature Discharge of Battery		
					CT3001	Implantable metal			A070504	Premature Discharge of Battery		
					CT3002	Implantable synthetic polymer			A070504	Premature Discharge of Battery		
					CT3006	Implantable inorganic material			A070504	Premature Discharge of Battery		
					CT314	Electrically powered			A070504	Premature Discharge of Battery		
					CT321	Long-term surgical invasive			A070504	Premature Discharge of Battery		
					CT384	Sensors and associated devices			A070504	Premature Discharge of Battery		
					CT449	Generators and associated devices			A070504	Premature Discharge of Battery		
					CT981	Single use			A070504	Premature Discharge of Battery		
					CT983	Surgical invasive			A070504	Premature Discharge of Battery		
					CT994	Sensors			A070504	Premature Discharge of Battery		
11765	2022		12913	Pacemaker, implantable	CT145	Cardiovascular devices	F23	Unexpected Medical Intervention	A01	Patient Device Interaction Problem		
									A0705	Battery Problem		
					CT1684	Cardiac pacemakers and associated devices	F23	Unexpected Medical Intervention	A01	Patient Device Interaction Problem		
									A0705	Battery Problem		
					CT1685	Cardiac pacemakers	F23	Unexpected Medical Intervention	A01	Patient Device Interaction Problem		
									A0705	Battery Problem		
					CT2406	Implantable	F23	Unexpected Medical Intervention	A01	Patient Device Interaction Problem		
									A0705	Battery Problem		
					CT2407	Active implantable	F23	Unexpected Medical Intervention	A01	Patient Device Interaction Problem		
									A0705	Battery Problem		
					CT257	Implantable cardiac pacemakers	F23	Unexpected Medical Intervention	A01	Patient Device Interaction Problem		
									A0705	Battery Problem		
					CT2727	Active implant devices/ancillaries	F23	Unexpected Medical Intervention	A01	Patient Device Interaction Problem		
									A0705	Battery Problem		
					CT2750	Implantable cardiac electrical stimulators and associated devices	F23	Unexpected Medical Intervention	A01	Patient Device Interaction Problem		
									A0705	Battery Problem		
					CT2866	Implantable cardiac electrical stimulators	F23	Unexpected Medical Intervention	A01	Patient Device Interaction Problem		
									A0705	Battery Problem		
					CT2985	Active	F23	Unexpected Medical Intervention	A01	Patient Device Interaction Problem		
									A0705	Battery Problem		
					CT314	Electrically powered	F23	Unexpected Medical Intervention	A01	Patient Device Interaction Problem		
									A0705	Battery Problem		
11766	2022	All other reportable incidents	47270	Cardiac resynchronization therapy implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E0601	Arrhythmia	A070912	Incorrect Interpretation of Signal		
									A071204	No Pacing		
							F23	Unexpected Medical Intervention	A070912	Incorrect Interpretation of Signal		
									A071204	No Pacing		
11767	2022		47265	Dual-chamber implantable pacemaker, rate-responsive	CT141	Electrodes/Leads and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem		
							F1901	Additional Surgery	A24	No Apparent Device Problem		
					CT145	Cardiovascular devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem		
							F1901	Additional Surgery	A24	No Apparent Device Problem		

					CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem		
							F1901	Additional Surgery	A24	No Apparent Device Problem		
					CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem		
							F1901	Additional Surgery	A24	No Apparent Device Problem		
					CT233	Surgical	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem		
							F1901	Additional Surgery	A24	No Apparent Device Problem		
					CT2406	Implantable	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem		
							F1901	Additional Surgery	A24	No Apparent Device Problem		
					CT2407	Active implantable	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem		
							F1901	Additional Surgery	A24	No Apparent Device Problem		
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem		
							F1901	Additional Surgery	A24	No Apparent Device Problem		
					CT2727	Active implant devices/ancillaries	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem		
							F1901	Additional Surgery	A24	No Apparent Device Problem		
					CT2750	Implantable cardiac electrical stimulators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem		
							F1901	Additional Surgery	A24	No Apparent Device Problem		
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem		
							F1901	Additional Surgery	A24	No Apparent Device Problem		
					CT2985	Active	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem		
							F1901	Additional Surgery	A24	No Apparent Device Problem		
					CT3001	Implantable metal	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem		
							F1901	Additional Surgery	A24	No Apparent Device Problem		
					CT3002	Implantable synthetic polymer	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem		
							F1901	Additional Surgery	A24	No Apparent Device Problem		
					CT3006	Implantable inorganic material	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem		
							F1901	Additional Surgery	A24	No Apparent Device Problem		
					CT314	Electrically powered	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem		
							F1901	Additional Surgery	A24	No Apparent Device Problem		
					CT321	Long-term surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem		
							F1901	Additional Surgery	A24	No Apparent Device Problem		
					CT449	Generators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem		
							F1901	Additional Surgery	A24	No Apparent Device Problem		
					CT981	Single use	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem		
							F1901	Additional Surgery	A24	No Apparent Device Problem		
					CT983	Surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem		
							F1901	Additional Surgery	A24	No Apparent Device Problem		
11768	2022		47265	Dual-chamber implantable pacemaker, rate-responsive	CT141	Electrodes/Leads and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	27/10/2022	27/10/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	27/10/2022	27/10/2022
					CT145	Cardiovascular devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	27/10/2022	27/10/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	27/10/2022	27/10/2022
					CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	27/10/2022	27/10/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	27/10/2022	27/10/2022
					CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	27/10/2022	27/10/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	27/10/2022	27/10/2022
					CT233	Surgical	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	27/10/2022	27/10/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	27/10/2022	27/10/2022
					CT2406	Implantable	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	27/10/2022	27/10/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	27/10/2022	27/10/2022
					CT2407	Active implantable	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	27/10/2022	27/10/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	27/10/2022	27/10/2022
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	27/10/2022	27/10/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	27/10/2022	27/10/2022
					CT2727	Active implant devices/ancillaries	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	27/10/2022	27/10/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	27/10/2022	27/10/2022
					CT2750	Implantable cardiac electrical stimulators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	27/10/2022	27/10/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	27/10/2022	27/10/2022
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	27/10/2022	27/10/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	27/10/2022	27/10/2022
					CT2985	Active	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	27/10/2022	27/10/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	27/10/2022	27/10/2022
					CT3001	Implantable metal	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	27/10/2022	27/10/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	27/10/2022	27/10/2022
					CT3002	Implantable synthetic polymer	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	27/10/2022	27/10/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	27/10/2022	27/10/2022
					CT3006	Implantable inorganic material	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	27/10/2022	27/10/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	27/10/2022	27/10/2022
					CT314	Electrically powered	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	27/10/2022	27/10/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	27/10/2022	27/10/2022
					CT321	Long-term surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	27/10/2022	27/10/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	27/10/2022	27/10/2022
					CT449	Generators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	27/10/2022	27/10/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	27/10/2022	27/10/2022
					CT981	Single use	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	27/10/2022	27/10/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	27/10/2022	27/10/2022
					CT983	Surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	27/10/2022	27/10/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	27/10/2022	27/10/2022
11769	2022		47265	Dual-chamber implantable pacemaker, rate-responsive	CT141	Electrodes/Leads and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	03/11/2022	03/11/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	03/11/2022	03/11/2022
					CT145	Cardiovascular devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	03/11/2022	03/11/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	03/11/2022	03/11/2022
					CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	03/11/2022	03/11/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	03/11/2022	03/11/2022
					CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	03/11/2022	03/11/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	03/11/2022	03/11/2022
					CT233	Surgical	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	03/11/2022	03/11/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	03/11/2022	03/11/2022
					CT2406	Implantable	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	03/11/2022	03/11/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	03/11/2022	03/11/2022

					CT2407	Active implantable	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	03/11/2022	03/11/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	03/11/2022	03/11/2022
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	03/11/2022	03/11/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	03/11/2022	03/11/2022
					CT2727	Active implant devices/ancillaries	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	03/11/2022	03/11/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	03/11/2022	03/11/2022
					CT2750	Implantable cardiac electrical stimulators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	03/11/2022	03/11/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	03/11/2022	03/11/2022
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	03/11/2022	03/11/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	03/11/2022	03/11/2022
					CT2985	Active	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	03/11/2022	03/11/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	03/11/2022	03/11/2022
					CT3001	Implantable metal	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	03/11/2022	03/11/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	03/11/2022	03/11/2022
					CT3002	Implantable synthetic polymer	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	03/11/2022	03/11/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	03/11/2022	03/11/2022
					CT3006	Implantable inorganic material	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	03/11/2022	03/11/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	03/11/2022	03/11/2022
					CT314	Electrically powered	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	03/11/2022	03/11/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	03/11/2022	03/11/2022
					CT321	Long-term surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	03/11/2022	03/11/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	03/11/2022	03/11/2022
					CT449	Generators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	03/11/2022	03/11/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	03/11/2022	03/11/2022
					CT981	Single use	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	03/11/2022	03/11/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	03/11/2022	03/11/2022
					CT983	Surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	03/11/2022	03/11/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	03/11/2022	03/11/2022
11770	2022		47265	Dual-chamber implantable pacemaker, rate-responsive	CT141	Electrodes/Leads and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	28/09/2022	28/09/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	28/09/2022	28/09/2022
					CT145	Cardiovascular devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	28/09/2022	28/09/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	28/09/2022	28/09/2022
					CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	28/09/2022	28/09/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	28/09/2022	28/09/2022
					CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	28/09/2022	28/09/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	28/09/2022	28/09/2022
					CT233	Surgical	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	28/09/2022	28/09/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	28/09/2022	28/09/2022
					CT2406	Implantable	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	28/09/2022	28/09/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	28/09/2022	28/09/2022
					CT2407	Active implantable	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	28/09/2022	28/09/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	28/09/2022	28/09/2022
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	28/09/2022	28/09/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	28/09/2022	28/09/2022
					CT2727	Active implant devices/ancillaries	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	28/09/2022	28/09/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	28/09/2022	28/09/2022
					CT2750	Implantable cardiac electrical stimulators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	28/09/2022	28/09/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	28/09/2022	28/09/2022
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	28/09/2022	28/09/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	28/09/2022	28/09/2022
					CT2985	Active	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	28/09/2022	28/09/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	28/09/2022	28/09/2022
					CT3001	Implantable metal	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	28/09/2022	28/09/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	28/09/2022	28/09/2022
					CT3002	Implantable synthetic polymer	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	28/09/2022	28/09/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	28/09/2022	28/09/2022
					CT3006	Implantable inorganic material	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	28/09/2022	28/09/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	28/09/2022	28/09/2022
					CT314	Electrically powered	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	28/09/2022	28/09/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	28/09/2022	28/09/2022
					CT321	Long-term surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	28/09/2022	28/09/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	28/09/2022	28/09/2022
					CT449	Generators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	28/09/2022	28/09/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	28/09/2022	28/09/2022
					CT981	Single use	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	28/09/2022	28/09/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	28/09/2022	28/09/2022
					CT983	Surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	28/09/2022	28/09/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	28/09/2022	28/09/2022
11771	2022		47265	Dual-chamber implantable pacemaker, rate-responsive	CT141	Electrodes/Leads and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	15/09/2022	15/09/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	15/09/2022	15/09/2022
					CT145	Cardiovascular devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	15/09/2022	15/09/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	15/09/2022	15/09/2022
					CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	15/09/2022	15/09/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	15/09/2022	15/09/2022
					CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	15/09/2022	15/09/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	15/09/2022	15/09/2022
					CT233	Surgical	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	15/09/2022	15/09/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	15/09/2022	15/09/2022
					CT2406	Implantable	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	15/09/2022	15/09/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	15/09/2022	15/09/2022
					CT2407	Active implantable	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	15/09/2022	15/09/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	15/09/2022	15/09/2022
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	15/09/2022	15/09/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	15/09/2022	15/09/2022
					CT2727	Active implant devices/ancillaries	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	15/09/2022	15/09/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	15/09/2022	15/09/2022
					CT2750	Implantable cardiac electrical stimulators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	15/09/2022	15/09/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	15/09/2022	15/09/2022

					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	15/09/2022	15/09/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	15/09/2022	15/09/2022
					CT2985	Active	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	15/09/2022	15/09/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	15/09/2022	15/09/2022
					CT3001	Implantable metal	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	15/09/2022	15/09/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	15/09/2022	15/09/2022
					CT3002	Implantable synthetic polymer	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	15/09/2022	15/09/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	15/09/2022	15/09/2022
					CT3006	Implantable inorganic material	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	15/09/2022	15/09/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	15/09/2022	15/09/2022
					CT314	Electrically powered	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	15/09/2022	15/09/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	15/09/2022	15/09/2022
					CT321	Long-term surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	15/09/2022	15/09/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	15/09/2022	15/09/2022
					CT449	Generators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	15/09/2022	15/09/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	15/09/2022	15/09/2022
					CT981	Single use	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	15/09/2022	15/09/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	15/09/2022	15/09/2022
					CT983	Surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	15/09/2022	15/09/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	15/09/2022	15/09/2022
11772	2022		47265	Dual-chamber implantable pacemaker, rate-responsive	CT141	Electrodes/Leads and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	10/10/2022	10/10/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	10/10/2022	10/10/2022
					CT145	Cardiovascular devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	10/10/2022	10/10/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	10/10/2022	10/10/2022
					CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	10/10/2022	10/10/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	10/10/2022	10/10/2022
					CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	10/10/2022	10/10/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	10/10/2022	10/10/2022
					CT233	Surgical	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	10/10/2022	10/10/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	10/10/2022	10/10/2022
					CT2406	Implantable	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	10/10/2022	10/10/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	10/10/2022	10/10/2022
					CT2407	Active implantable	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	10/10/2022	10/10/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	10/10/2022	10/10/2022
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	10/10/2022	10/10/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	10/10/2022	10/10/2022
					CT2727	Active implant devices/ancillaries	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	10/10/2022	10/10/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	10/10/2022	10/10/2022
					CT2750	Implantable cardiac electrical stimulators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	10/10/2022	10/10/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	10/10/2022	10/10/2022
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	10/10/2022	10/10/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	10/10/2022	10/10/2022
					CT2985	Active	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	10/10/2022	10/10/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	10/10/2022	10/10/2022
					CT3001	Implantable metal	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	10/10/2022	10/10/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	10/10/2022	10/10/2022
					CT3002	Implantable synthetic polymer	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	10/10/2022	10/10/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	10/10/2022	10/10/2022
					CT3006	Implantable inorganic material	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	10/10/2022	10/10/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	10/10/2022	10/10/2022
					CT314	Electrically powered	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	10/10/2022	10/10/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	10/10/2022	10/10/2022
					CT321	Long-term surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	10/10/2022	10/10/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	10/10/2022	10/10/2022
					CT449	Generators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	10/10/2022	10/10/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	10/10/2022	10/10/2022
					CT981	Single use	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	10/10/2022	10/10/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	10/10/2022	10/10/2022
					CT983	Surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	10/10/2022	10/10/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	10/10/2022	10/10/2022
11773	2022		47265	Dual-chamber implantable pacemaker, rate-responsive	CT141	Electrodes/Leads and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	22/11/2022	22/11/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	22/11/2022	22/11/2022
					CT145	Cardiovascular devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	22/11/2022	22/11/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	22/11/2022	22/11/2022
					CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	22/11/2022	22/11/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	22/11/2022	22/11/2022
					CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	22/11/2022	22/11/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	22/11/2022	22/11/2022
					CT233	Surgical	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	22/11/2022	22/11/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	22/11/2022	22/11/2022
					CT2406	Implantable	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	22/11/2022	22/11/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	22/11/2022	22/11/2022
					CT2407	Active implantable	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	22/11/2022	22/11/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	22/11/2022	22/11/2022
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	22/11/2022	22/11/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	22/11/2022	22/11/2022
					CT2727	Active implant devices/ancillaries	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	22/11/2022	22/11/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	22/11/2022	22/11/2022
					CT2750	Implantable cardiac electrical stimulators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	22/11/2022	22/11/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	22/11/2022	22/11/2022
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	22/11/2022	22/11/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	22/11/2022	22/11/2022
					CT2985	Active	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	22/11/2022	22/11/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	22/11/2022	22/11/2022
					CT3001	Implantable metal	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	22/11/2022	22/11/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	22/11/2022	22/11/2022
					CT3002	Implantable synthetic polymer	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	22/11/2022	22/11/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	22/11/2022	22/11/2022

					CT3006	Implantable inorganic material	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	22/11/2022	22/11/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	22/11/2022	22/11/2022
					CT314	Electrically powered	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	22/11/2022	22/11/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	22/11/2022	22/11/2022
					CT321	Long-term surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	22/11/2022	22/11/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	22/11/2022	22/11/2022
					CT449	Generators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	22/11/2022	22/11/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	22/11/2022	22/11/2022
					CT981	Single use	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	22/11/2022	22/11/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	22/11/2022	22/11/2022
					CT983	Surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	22/11/2022	22/11/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	22/11/2022	22/11/2022
11774	2022		47265	Dual-chamber implantable pacemaker, rate-responsive	CT141	Electrodes/Leads and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	13/10/2022	13/10/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	13/10/2022	13/10/2022
					CT145	Cardiovascular devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	13/10/2022	13/10/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	13/10/2022	13/10/2022
					CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	13/10/2022	13/10/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	13/10/2022	13/10/2022
					CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	13/10/2022	13/10/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	13/10/2022	13/10/2022
					CT233	Surgical	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	13/10/2022	13/10/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	13/10/2022	13/10/2022
					CT2406	Implantable	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	13/10/2022	13/10/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	13/10/2022	13/10/2022
					CT2407	Active implantable	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	13/10/2022	13/10/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	13/10/2022	13/10/2022
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	13/10/2022	13/10/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	13/10/2022	13/10/2022
					CT2727	Active implant devices/ancillaries	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	13/10/2022	13/10/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	13/10/2022	13/10/2022
					CT2750	Implantable cardiac electrical stimulators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	13/10/2022	13/10/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	13/10/2022	13/10/2022
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	13/10/2022	13/10/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	13/10/2022	13/10/2022
					CT2985	Active	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	13/10/2022	13/10/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	13/10/2022	13/10/2022
					CT3001	Implantable metal	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	13/10/2022	13/10/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	13/10/2022	13/10/2022
					CT3002	Implantable synthetic polymer	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	13/10/2022	13/10/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	13/10/2022	13/10/2022
					CT3006	Implantable inorganic material	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	13/10/2022	13/10/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	13/10/2022	13/10/2022
					CT314	Electrically powered	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	13/10/2022	13/10/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	13/10/2022	13/10/2022
					CT321	Long-term surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	13/10/2022	13/10/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	13/10/2022	13/10/2022
					CT449	Generators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	13/10/2022	13/10/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	13/10/2022	13/10/2022
					CT981	Single use	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	13/10/2022	13/10/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	13/10/2022	13/10/2022
					CT983	Surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	13/10/2022	13/10/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	13/10/2022	13/10/2022
11775	2022	All other reportable incidents	47265	Dual-chamber implantable pacemaker, rate-responsive	CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	21/11/2022	21/11/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	21/11/2022	21/11/2022
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	21/11/2022	21/11/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	21/11/2022	21/11/2022
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	21/11/2022	21/11/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	21/11/2022	21/11/2022
11776	2022		47265	Dual-chamber implantable pacemaker, rate-responsive	CT141	Electrodes/Leads and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	08/09/2022	08/09/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	08/09/2022	08/09/2022
					CT145	Cardiovascular devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	08/09/2022	08/09/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	08/09/2022	08/09/2022
					CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	08/09/2022	08/09/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	08/09/2022	08/09/2022
					CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	08/09/2022	08/09/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	08/09/2022	08/09/2022
					CT233	Surgical	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	08/09/2022	08/09/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	08/09/2022	08/09/2022
					CT2406	Implantable	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	08/09/2022	08/09/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	08/09/2022	08/09/2022
					CT2407	Active implantable	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	08/09/2022	08/09/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	08/09/2022	08/09/2022
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	08/09/2022	08/09/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	08/09/2022	08/09/2022
					CT2727	Active implant devices/ancillaries	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	08/09/2022	08/09/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	08/09/2022	08/09/2022
					CT2750	Implantable cardiac electrical stimulators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	08/09/2022	08/09/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	08/09/2022	08/09/2022
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	08/09/2022	08/09/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	08/09/2022	08/09/2022
					CT2985	Active	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	08/09/2022	08/09/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	08/09/2022	08/09/2022
					CT3001	Implantable metal	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	08/09/2022	08/09/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	08/09/2022	08/09/2022
					CT3002	Implantable synthetic polymer	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	08/09/2022	08/09/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	08/09/2022	08/09/2022
					CT3006	Implantable inorganic material	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	08/09/2022	08/09/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	08/09/2022	08/09/2022

					CT314	Electrically powered	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	08/09/2022	08/09/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	08/09/2022	08/09/2022
					CT321	Long-term surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	08/09/2022	08/09/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	08/09/2022	08/09/2022
					CT449	Generators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	08/09/2022	08/09/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	08/09/2022	08/09/2022
					CT981	Single use	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	08/09/2022	08/09/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	08/09/2022	08/09/2022
					CT983	Surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	08/09/2022	08/09/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	08/09/2022	08/09/2022
11777	2022	All other reportable incidents	36241	Pacemaker/defibrillator lead	CT1684	Cardiac pacemakers and associated devices	E0112	Dizziness	A070101	Failure to Capture		
									A070909	Oversensing		
									A0712	Pacing Problem		
									A071201	Failure to Convert Rhythm		
							E011903	Syncope/Fainting	A070101	Failure to Capture		
									A070909	Oversensing		
									A0712	Pacing Problem		
									A071201	Failure to Convert Rhythm		
							F05	Delay to Treatment/ Therapy	A070101	Failure to Capture		
									A070909	Oversensing		
									A0712	Pacing Problem		
									A071201	Failure to Convert Rhythm		
							F08	Hospitalization or Prolonged Hospitalization	A070101	Failure to Capture		
									A070909	Oversensing		
									A0712	Pacing Problem		
									A071201	Failure to Convert Rhythm		
							F1001	Absence of Treatment	A070101	Failure to Capture		
									A070909	Oversensing		
									A0712	Pacing Problem		
									A071201	Failure to Convert Rhythm		
							F2203	Imaging Required	A070101	Failure to Capture		
									A070909	Oversensing		
									A0712	Pacing Problem		
									A071201	Failure to Convert Rhythm		
11778	2022	All other reportable incidents	47265	Dual-chamber implantable pacemaker, rate-responsive	CT141	Electrodes/Leads and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	05/12/2022	05/12/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	05/12/2022	05/12/2022
					CT145	Cardiovascular devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	05/12/2022	05/12/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	05/12/2022	05/12/2022
					CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	05/12/2022	05/12/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	05/12/2022	05/12/2022
					CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	05/12/2022	05/12/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	05/12/2022	05/12/2022
					CT233	Surgical	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	05/12/2022	05/12/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	05/12/2022	05/12/2022
					CT2406	Implantable	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	05/12/2022	05/12/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	05/12/2022	05/12/2022
					CT2407	Active implantable	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	05/12/2022	05/12/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	05/12/2022	05/12/2022
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	05/12/2022	05/12/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	05/12/2022	05/12/2022
					CT2727	Active implant devices/ancillaries	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	05/12/2022	05/12/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	05/12/2022	05/12/2022
					CT2750	Implantable cardiac electrical stimulators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	05/12/2022	05/12/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	05/12/2022	05/12/2022
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	05/12/2022	05/12/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	05/12/2022	05/12/2022
					CT2985	Active	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	05/12/2022	05/12/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	05/12/2022	05/12/2022
					CT3001	Implantable metal	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	05/12/2022	05/12/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	05/12/2022	05/12/2022
					CT3002	Implantable synthetic polymer	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	05/12/2022	05/12/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	05/12/2022	05/12/2022
					CT3006	Implantable inorganic material	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	05/12/2022	05/12/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	05/12/2022	05/12/2022
					CT314	Electrically powered	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	05/12/2022	05/12/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	05/12/2022	05/12/2022
					CT321	Long-term surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	05/12/2022	05/12/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	05/12/2022	05/12/2022
					CT449	Generators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	05/12/2022	05/12/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	05/12/2022	05/12/2022
					CT981	Single use	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	05/12/2022	05/12/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	05/12/2022	05/12/2022
					CT983	Surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	05/12/2022	05/12/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	05/12/2022	05/12/2022
11779	2022		47265	Dual-chamber implantable pacemaker, rate-responsive	CT141	Electrodes/Leads and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	13/10/2022	13/10/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	13/10/2022	13/10/2022
					CT145	Cardiovascular devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	13/10/2022	13/10/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	13/10/2022	13/10/2022
					CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	13/10/2022	13/10/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	13/10/2022	13/10/2022
					CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	13/10/2022	13/10/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	13/10/2022	13/10/2022
					CT233	Surgical	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	13/10/2022	13/10/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	13/10/2022	13/10/2022
					CT2406	Implantable	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	13/10/2022	13/10/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	13/10/2022	13/10/2022
					CT2407	Active implantable	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	13/10/2022	13/10/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	13/10/2022	13/10/2022

					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	13/10/2022	13/10/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	13/10/2022	13/10/2022
					CT2727	Active implant devices/ancillaries	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	13/10/2022	13/10/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	13/10/2022	13/10/2022
					CT2750	Implantable cardiac electrical stimulators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	13/10/2022	13/10/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	13/10/2022	13/10/2022
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	13/10/2022	13/10/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	13/10/2022	13/10/2022
					CT2985	Active	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	13/10/2022	13/10/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	13/10/2022	13/10/2022
					CT3001	Implantable metal	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	13/10/2022	13/10/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	13/10/2022	13/10/2022
					CT3002	Implantable synthetic polymer	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	13/10/2022	13/10/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	13/10/2022	13/10/2022
					CT3006	Implantable inorganic material	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	13/10/2022	13/10/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	13/10/2022	13/10/2022
					CT314	Electrically powered	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	13/10/2022	13/10/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	13/10/2022	13/10/2022
					CT321	Long-term surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	13/10/2022	13/10/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	13/10/2022	13/10/2022
					CT449	Generators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	13/10/2022	13/10/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	13/10/2022	13/10/2022
					CT981	Single use	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	13/10/2022	13/10/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	13/10/2022	13/10/2022
					CT983	Surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	13/10/2022	13/10/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	13/10/2022	13/10/2022
11780	2022	All other reportable incidents	47265	Dual-chamber implantable pacemaker, rate-responsive	CT141	Electrodes/Leads and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	30/11/2022	30/11/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	30/11/2022	30/11/2022
					CT145	Cardiovascular devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	30/11/2022	30/11/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	30/11/2022	30/11/2022
					CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	30/11/2022	30/11/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	30/11/2022	30/11/2022
					CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	30/11/2022	30/11/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	30/11/2022	30/11/2022
					CT233	Surgical	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	30/11/2022	30/11/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	30/11/2022	30/11/2022
					CT2406	Implantable	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	30/11/2022	30/11/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	30/11/2022	30/11/2022
					CT2407	Active implantable	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	30/11/2022	30/11/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	30/11/2022	30/11/2022
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	30/11/2022	30/11/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	30/11/2022	30/11/2022
					CT2727	Active implant devices/ancillaries	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	30/11/2022	30/11/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	30/11/2022	30/11/2022
					CT2750	Implantable cardiac electrical stimulators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	30/11/2022	30/11/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	30/11/2022	30/11/2022
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	30/11/2022	30/11/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	30/11/2022	30/11/2022
					CT2985	Active	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	30/11/2022	30/11/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	30/11/2022	30/11/2022
					CT3001	Implantable metal	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	30/11/2022	30/11/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	30/11/2022	30/11/2022
					CT3002	Implantable synthetic polymer	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	30/11/2022	30/11/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	30/11/2022	30/11/2022
					CT3006	Implantable inorganic material	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	30/11/2022	30/11/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	30/11/2022	30/11/2022
					CT314	Electrically powered	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	30/11/2022	30/11/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	30/11/2022	30/11/2022
					CT321	Long-term surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	30/11/2022	30/11/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	30/11/2022	30/11/2022
					CT449	Generators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	30/11/2022	30/11/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	30/11/2022	30/11/2022
					CT981	Single use	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	30/11/2022	30/11/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	30/11/2022	30/11/2022
					CT983	Surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	30/11/2022	30/11/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	30/11/2022	30/11/2022
11781	2022		47265	Dual-chamber implantable pacemaker, rate-responsive	CT141	Electrodes/Leads and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	28/10/2022	28/10/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	28/10/2022	28/10/2022
					CT145	Cardiovascular devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	28/10/2022	28/10/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	28/10/2022	28/10/2022
					CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	28/10/2022	28/10/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	28/10/2022	28/10/2022
					CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	28/10/2022	28/10/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	28/10/2022	28/10/2022
					CT233	Surgical	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	28/10/2022	28/10/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	28/10/2022	28/10/2022
					CT2406	Implantable	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	28/10/2022	28/10/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	28/10/2022	28/10/2022
					CT2407	Active implantable	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	28/10/2022	28/10/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	28/10/2022	28/10/2022
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	28/10/2022	28/10/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	28/10/2022	28/10/2022
					CT2727	Active implant devices/ancillaries	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	28/10/2022	28/10/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	28/10/2022	28/10/2022
					CT2750	Implantable cardiac electrical stimulators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	28/10/2022	28/10/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	28/10/2022	28/10/2022
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	28/10/2022	28/10/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	28/10/2022	28/10/2022

					CT2985	Active	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	28/10/2022	28/10/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	28/10/2022	28/10/2022
					CT3001	Implantable metal	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	28/10/2022	28/10/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	28/10/2022	28/10/2022
					CT3002	Implantable synthetic polymer	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	28/10/2022	28/10/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	28/10/2022	28/10/2022
					CT3006	Implantable inorganic material	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	28/10/2022	28/10/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	28/10/2022	28/10/2022
					CT314	Electrically powered	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	28/10/2022	28/10/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	28/10/2022	28/10/2022
					CT321	Long-term surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	28/10/2022	28/10/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	28/10/2022	28/10/2022
					CT449	Generators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	28/10/2022	28/10/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	28/10/2022	28/10/2022
					CT981	Single use	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	28/10/2022	28/10/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	28/10/2022	28/10/2022
					CT983	Surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	28/10/2022	28/10/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	28/10/2022	28/10/2022
11782	2022		47265	Dual-chamber implantable pacemaker, rate-responsive	CT141	Electrodes/Leads and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	27/10/2022	27/10/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	27/10/2022	27/10/2022
					CT145	Cardiovascular devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	27/10/2022	27/10/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	27/10/2022	27/10/2022
					CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	27/10/2022	27/10/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	27/10/2022	27/10/2022
					CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	27/10/2022	27/10/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	27/10/2022	27/10/2022
					CT233	Surgical	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	27/10/2022	27/10/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	27/10/2022	27/10/2022
					CT2406	Implantable	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	27/10/2022	27/10/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	27/10/2022	27/10/2022
					CT2407	Active implantable	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	27/10/2022	27/10/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	27/10/2022	27/10/2022
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	27/10/2022	27/10/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	27/10/2022	27/10/2022
					CT2727	Active implant devices/ancillaries	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	27/10/2022	27/10/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	27/10/2022	27/10/2022
					CT2750	Implantable cardiac electrical stimulators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	27/10/2022	27/10/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	27/10/2022	27/10/2022
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	27/10/2022	27/10/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	27/10/2022	27/10/2022
					CT2985	Active	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	27/10/2022	27/10/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	27/10/2022	27/10/2022
					CT3001	Implantable metal	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	27/10/2022	27/10/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	27/10/2022	27/10/2022
					CT3002	Implantable synthetic polymer	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	27/10/2022	27/10/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	27/10/2022	27/10/2022
					CT3006	Implantable inorganic material	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	27/10/2022	27/10/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	27/10/2022	27/10/2022
					CT314	Electrically powered	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	27/10/2022	27/10/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	27/10/2022	27/10/2022
					CT321	Long-term surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	27/10/2022	27/10/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	27/10/2022	27/10/2022
					CT449	Generators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	27/10/2022	27/10/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	27/10/2022	27/10/2022
					CT981	Single use	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	27/10/2022	27/10/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	27/10/2022	27/10/2022
					CT983	Surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	27/10/2022	27/10/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	27/10/2022	27/10/2022
11783	2022	All other reportable incidents	37265	Dual-chamber implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A070603	Failure to Charge	24/02/2022	24/02/2022
							F1901	Additional Surgery	A070603	Failure to Charge	24/02/2022	24/02/2022
11784	2019	All other reportable incidents	35852	Single-chamber implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	19/12/2018	
11785	2019	All other reportable incidents	35223	Endocardial/interventricular septal pacing lead	CT1684	Cardiac pacemakers and associated devices	F1905	Device Revision or Replacement	A0709	Device Sensing Problem		
11786	2019	All other reportable incidents	35852	Single-chamber implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A0709	Device Sensing Problem		
11787	2020	All other reportable incidents	47265	Dual-chamber implantable pacemaker, rate-responsive	CT141	Electrodes/Leads and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	07/07/2020	07/07/2020
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery	07/07/2020	07/07/2020
					CT145	Cardiovascular devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	07/07/2020	07/07/2020
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery	07/07/2020	07/07/2020
					CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	07/07/2020	07/07/2020
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery	07/07/2020	07/07/2020
					CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	07/07/2020	07/07/2020
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery	07/07/2020	07/07/2020
					CT233	Surgical	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	07/07/2020	07/07/2020
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery	07/07/2020	07/07/2020
					CT2406	Implantable	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	07/07/2020	07/07/2020
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery	07/07/2020	07/07/2020
					CT2407	Active implantable	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	07/07/2020	07/07/2020
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery	07/07/2020	07/07/2020
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	07/07/2020	07/07/2020
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery	07/07/2020	07/07/2020
					CT2727	Active implant devices/ancillaries	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	07/07/2020	07/07/2020
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery	07/07/2020	07/07/2020
					CT2750	Implantable cardiac electrical stimulators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	07/07/2020	07/07/2020
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery	07/07/2020	07/07/2020
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	07/07/2020	07/07/2020
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery	07/07/2020	07/07/2020
					CT2985	Active	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	07/07/2020	07/07/2020
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery	07/07/2020	07/07/2020

					CT3001	Implantable metal	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	07/07/2020	07/07/2020
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery	07/07/2020	07/07/2020
					CT3002	Implantable synthetic polymer	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	07/07/2020	07/07/2020
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery	07/07/2020	07/07/2020
					CT3006	Implantable inorganic material	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	07/07/2020	07/07/2020
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery	07/07/2020	07/07/2020
					CT314	Electrically powered	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	07/07/2020	07/07/2020
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery	07/07/2020	07/07/2020
					CT321	Long-term surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	07/07/2020	07/07/2020
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery	07/07/2020	07/07/2020
					CT449	Generators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	07/07/2020	07/07/2020
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery	07/07/2020	07/07/2020
					CT981	Single use	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	07/07/2020	07/07/2020
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery	07/07/2020	07/07/2020
					CT983	Surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	07/07/2020	07/07/2020
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery	07/07/2020	07/07/2020
11788	2020	All other reportable incidents	47265	Dual-chamber implantable pacemaker, rate-responsive	CT141	Electrodes/Leads and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A23	Use of Device Problem		
							F26	No Health Consequences or Impact	A23	Use of Device Problem		
					CT145	Cardiovascular devices	E2403	No Clinical Signs, Symptoms or Conditions	A23	Use of Device Problem		
							F26	No Health Consequences or Impact	A23	Use of Device Problem		
					CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A23	Use of Device Problem		
							F26	No Health Consequences or Impact	A23	Use of Device Problem		
					CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A23	Use of Device Problem		
							F26	No Health Consequences or Impact	A23	Use of Device Problem		
					CT233	Surgical	E2403	No Clinical Signs, Symptoms or Conditions	A23	Use of Device Problem		
							F26	No Health Consequences or Impact	A23	Use of Device Problem		
					CT2406	Implantable	E2403	No Clinical Signs, Symptoms or Conditions	A23	Use of Device Problem		
							F26	No Health Consequences or Impact	A23	Use of Device Problem		
					CT2407	Active implantable	E2403	No Clinical Signs, Symptoms or Conditions	A23	Use of Device Problem		
							F26	No Health Consequences or Impact	A23	Use of Device Problem		
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A23	Use of Device Problem		
							F26	No Health Consequences or Impact	A23	Use of Device Problem		
					CT2727	Active implant devices/ancillaries	E2403	No Clinical Signs, Symptoms or Conditions	A23	Use of Device Problem		
							F26	No Health Consequences or Impact	A23	Use of Device Problem		
					CT2750	Implantable cardiac electrical stimulators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A23	Use of Device Problem		
							F26	No Health Consequences or Impact	A23	Use of Device Problem		
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A23	Use of Device Problem		
							F26	No Health Consequences or Impact	A23	Use of Device Problem		
					CT2985	Active	E2403	No Clinical Signs, Symptoms or Conditions	A23	Use of Device Problem		
							F26	No Health Consequences or Impact	A23	Use of Device Problem		
					CT3001	Implantable metal	E2403	No Clinical Signs, Symptoms or Conditions	A23	Use of Device Problem		
							F26	No Health Consequences or Impact	A23	Use of Device Problem		
					CT3002	Implantable synthetic polymer	E2403	No Clinical Signs, Symptoms or Conditions	A23	Use of Device Problem		
							F26	No Health Consequences or Impact	A23	Use of Device Problem		
					CT3006	Implantable inorganic material	E2403	No Clinical Signs, Symptoms or Conditions	A23	Use of Device Problem		
							F26	No Health Consequences or Impact	A23	Use of Device Problem		
					CT314	Electrically powered	E2403	No Clinical Signs, Symptoms or Conditions	A23	Use of Device Problem		
							F26	No Health Consequences or Impact	A23	Use of Device Problem		
					CT321	Long-term surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A23	Use of Device Problem		
							F26	No Health Consequences or Impact	A23	Use of Device Problem		
					CT449	Generators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A23	Use of Device Problem		
							F26	No Health Consequences or Impact	A23	Use of Device Problem		
					CT981	Single use	E2403	No Clinical Signs, Symptoms or Conditions	A23	Use of Device Problem		
							F26	No Health Consequences or Impact	A23	Use of Device Problem		
					CT983	Surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A23	Use of Device Problem		
							F26	No Health Consequences or Impact	A23	Use of Device Problem		
11789	2020	All other reportable incidents	17882	Physiologic-monitoring defibrillation system	CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A07	Electrical /Electronic Property Problem		
							F1001	Absence of Treatment	A07	Electrical /Electronic Property Problem		
							F26	No Health Consequences or Impact	A07	Electrical /Electronic Property Problem		
11790	2019	All other reportable incidents	37265	Dual-chamber implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A072201	High Impedance		
11791	2020	All other reportable incidents	36241	Pacemaker/defibrillator lead	CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A0203	Defective Device		
									A12	Connection Problem		
							F1905	Device Revision or Replacement	A0203	Defective Device		
									A12	Connection Problem		
11792	2020	All other reportable incidents	37265	Dual-chamber implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	20/05/2020	20/05/2020
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	20/05/2020	20/05/2020
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery	20/05/2020	20/05/2020
11793	2020	All other reportable incidents	35223	Endocardial/interventricular septal pacing lead	CT1684	Cardiac pacemakers and associated devices	E2401	Insufficient Information	A070910	Undersensing	14/07/2020	14/07/2020
							F1905	Device Revision or Replacement	A070910	Undersensing	14/07/2020	14/07/2020
11794	2019	All other reportable incidents	36241	Pacemaker/defibrillator lead	CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070909	Oversensing		
									A090801	Signal Artifact/Noise		
							F26	No Health Consequences or Impact	A070909	Oversensing		
									A090801	Signal Artifact/Noise		
11795	2020	All other reportable incidents	37265	Dual-chamber implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
11796	2019	All other reportable incidents	36241	Pacemaker/defibrillator lead	CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A	Annex A		
11797	2021	All other reportable incidents	47265	Dual-chamber implantable pacemaker, rate-responsive	CT141	Electrodes/Leads and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT145	Cardiovascular devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT233	Surgical	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT2406	Implantable	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		

					CT2407	Active implantable	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT2727	Active implant devices/ancillaries	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT2750	Implantable cardiac electrical stimulators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT2985	Active	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT3001	Implantable metal	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT3002	Implantable synthetic polymer	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT3006	Implantable inorganic material	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT314	Electrically powered	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT321	Long-term surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT449	Generators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT981	Single use	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT983	Surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
11798	2019	All other reportable incidents	47265	Dual-chamber implantable pacemaker, rate-responsive	CT141	Electrodes/Leads and associated devices	F1905	Device Revision or Replacement	A0721	Circuit Failure	01/04/2019	
					CT145	Cardiovascular devices	F1905	Device Revision or Replacement	A0721	Circuit Failure	01/04/2019	
					CT1684	Cardiac pacemakers and associated devices	F1905	Device Revision or Replacement	A0721	Circuit Failure	01/04/2019	
					CT1685	Cardiac pacemakers	F1905	Device Revision or Replacement	A0721	Circuit Failure	01/04/2019	
					CT233	Surgical	F1905	Device Revision or Replacement	A0721	Circuit Failure	01/04/2019	
					CT2406	Implantable	F1905	Device Revision or Replacement	A0721	Circuit Failure	01/04/2019	
					CT2407	Active implantable	F1905	Device Revision or Replacement	A0721	Circuit Failure	01/04/2019	
					CT257	Implantable cardiac pacemakers	F1905	Device Revision or Replacement	A0721	Circuit Failure	01/04/2019	
					CT2727	Active implant devices/ancillaries	F1905	Device Revision or Replacement	A0721	Circuit Failure	01/04/2019	
					CT2750	Implantable cardiac electrical stimulators and associated devices	F1905	Device Revision or Replacement	A0721	Circuit Failure	01/04/2019	
					CT2866	Implantable cardiac electrical stimulators	F1905	Device Revision or Replacement	A0721	Circuit Failure	01/04/2019	
					CT2985	Active	F1905	Device Revision or Replacement	A0721	Circuit Failure	01/04/2019	
					CT3001	Implantable metal	F1905	Device Revision or Replacement	A0721	Circuit Failure	01/04/2019	
					CT3002	Implantable synthetic polymer	F1905	Device Revision or Replacement	A0721	Circuit Failure	01/04/2019	
					CT3006	Implantable inorganic material	F1905	Device Revision or Replacement	A0721	Circuit Failure	01/04/2019	
					CT314	Electrically powered	F1905	Device Revision or Replacement	A0721	Circuit Failure	01/04/2019	
					CT321	Long-term surgical invasive	F1905	Device Revision or Replacement	A0721	Circuit Failure	01/04/2019	
					CT449	Generators and associated devices	F1905	Device Revision or Replacement	A0721	Circuit Failure	01/04/2019	
					CT981	Single use	F1905	Device Revision or Replacement	A0721	Circuit Failure	01/04/2019	
					CT983	Surgical invasive	F1905	Device Revision or Replacement	A0721	Circuit Failure	01/04/2019	
11799	2019	All other reportable incidents	47263	Cardiac resynchronization therapy implantable pacemaker	CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
					CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
11800	2019	All other reportable incidents	47265	Dual-chamber implantable pacemaker, rate-responsive	CT141	Electrodes/Leads and associated devices	F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	01/11/2019	01/11/2019
					CT145	Cardiovascular devices	F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	01/11/2019	01/11/2019
					CT1684	Cardiac pacemakers and associated devices	F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	01/11/2019	01/11/2019
					CT1685	Cardiac pacemakers	F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	01/11/2019	01/11/2019
					CT233	Surgical	F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	01/11/2019	01/11/2019
					CT2406	Implantable	F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	01/11/2019	01/11/2019
					CT2407	Active implantable	F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	01/11/2019	01/11/2019
					CT257	Implantable cardiac pacemakers	F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	01/11/2019	01/11/2019
					CT2727	Active implant devices/ancillaries	F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	01/11/2019	01/11/2019
					CT2750	Implantable cardiac electrical stimulators and associated devices	F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	01/11/2019	01/11/2019
					CT2866	Implantable cardiac electrical stimulators	F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	01/11/2019	01/11/2019
					CT2985	Active	F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	01/11/2019	01/11/2019
					CT3001	Implantable metal	F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	01/11/2019	01/11/2019
					CT3002	Implantable synthetic polymer	F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	01/11/2019	01/11/2019
					CT3006	Implantable inorganic material	F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	01/11/2019	01/11/2019
					CT314	Electrically powered	F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	01/11/2019	01/11/2019
					CT321	Long-term surgical invasive	F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	01/11/2019	01/11/2019
					CT449	Generators and associated devices	F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	01/11/2019	01/11/2019
					CT981	Single use	F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	01/11/2019	01/11/2019
					CT983	Surgical invasive	F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	01/11/2019	01/11/2019
11801	2020	All other reportable incidents	47270	Cardiac resynchronization therapy implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	27/07/2020	27/07/2020
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	27/07/2020	27/07/2020
							F12	Serious Injury/ Illness/ Impairment	A070504	Premature Discharge of Battery	27/07/2020	27/07/2020
							F1903	Device Explantation	A070504	Premature Discharge of Battery	27/07/2020	27/07/2020
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	27/07/2020	27/07/2020
11802	2022		47265	Dual-chamber implantable pacemaker, rate-responsive	CT141	Electrodes/Leads and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem		
							F1901	Additional Surgery	A24	No Apparent Device Problem		
					CT145	Cardiovascular devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem		
							F1901	Additional Surgery	A24	No Apparent Device Problem		
					CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem		
							F1901	Additional Surgery	A24	No Apparent Device Problem		
					CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem		
							F1901	Additional Surgery	A24	No Apparent Device Problem		
					CT233	Surgical	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem		

							F1901	Additional Surgery	A24	No Apparent Device Problem		
					CT2406	Implantable	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem		
							F1901	Additional Surgery	A24	No Apparent Device Problem		
					CT2407	Active implantable	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem		
							F1901	Additional Surgery	A24	No Apparent Device Problem		
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem		
							F1901	Additional Surgery	A24	No Apparent Device Problem		
					CT2727	Active implant devices/ancillaries	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem		
							F1901	Additional Surgery	A24	No Apparent Device Problem		
					CT2750	Implantable cardiac electrical stimulators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem		
							F1901	Additional Surgery	A24	No Apparent Device Problem		
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem		
							F1901	Additional Surgery	A24	No Apparent Device Problem		
					CT2985	Active	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem		
							F1901	Additional Surgery	A24	No Apparent Device Problem		
					CT3001	Implantable metal	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem		
							F1901	Additional Surgery	A24	No Apparent Device Problem		
					CT3002	Implantable synthetic polymer	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem		
							F1901	Additional Surgery	A24	No Apparent Device Problem		
					CT3006	Implantable inorganic material	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem		
							F1901	Additional Surgery	A24	No Apparent Device Problem		
					CT314	Electrically powered	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem		
							F1901	Additional Surgery	A24	No Apparent Device Problem		
					CT321	Long-term surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem		
							F1901	Additional Surgery	A24	No Apparent Device Problem		
					CT449	Generators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem		
							F1901	Additional Surgery	A24	No Apparent Device Problem		
					CT981	Single use	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem		
							F1901	Additional Surgery	A24	No Apparent Device Problem		
					CT983	Surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem		
							F1901	Additional Surgery	A24	No Apparent Device Problem		
11803	2019	All other reportable incidents	35852	Single-chamber implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A070909	Oversensing		
									A071302	Inappropriate/Inadequate Shock/Stimulation		
11804	2020	All other reportable incidents	35223	Endocardial/interventricular septal pacing lead	CT1684	Cardiac pacemakers and associated devices	E0601	Arrhythmia	A070909	Oversensing	17/08/2020	17/08/2020
									A071204	No Pacing	17/08/2020	17/08/2020
									A090801	Signal Artifact/Noise	17/08/2020	17/08/2020
							F1905	Device Revision or Replacement	A070909	Oversensing	17/08/2020	17/08/2020
									A071204	No Pacing	17/08/2020	17/08/2020
									A090801	Signal Artifact/Noise	17/08/2020	17/08/2020
							F1908	Prolonged Surgery	A070909	Oversensing	17/08/2020	17/08/2020
									A071204	No Pacing	17/08/2020	17/08/2020
									A090801	Signal Artifact/Noise	17/08/2020	17/08/2020
11805	2021	All other reportable incidents	35852	Single-chamber implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E2104	Electric Shock	A070909	Oversensing		
									A071302	Inappropriate/Inadequate Shock/Stimulation		
									A090801	Signal Artifact/Noise		
							F13	Misdiagnosis/ Misclassification	A070909	Oversensing		
									A071302	Inappropriate/Inadequate Shock/Stimulation		
									A090801	Signal Artifact/Noise		
11806	2020	All other reportable incidents	36241	Pacemaker/defibrillator lead	CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A072201	High Impedance		
							F26	No Health Consequences or Impact	A072201	High Impedance		
11807	2020	All other reportable incidents	47265	Dual-chamber implantable pacemaker, rate-responsive	CT141	Electrodes/Leads and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	05/09/2020	05/09/2020
							F1901	Additional Surgery	A070504	Premature Discharge of Battery	05/09/2020	05/09/2020
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	05/09/2020	05/09/2020
					CT145	Cardiovascular devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	05/09/2020	05/09/2020
							F1901	Additional Surgery	A070504	Premature Discharge of Battery	05/09/2020	05/09/2020
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	05/09/2020	05/09/2020
					CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	05/09/2020	05/09/2020
							F1901	Additional Surgery	A070504	Premature Discharge of Battery	05/09/2020	05/09/2020
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	05/09/2020	05/09/2020
					CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	05/09/2020	05/09/2020
							F1901	Additional Surgery	A070504	Premature Discharge of Battery	05/09/2020	05/09/2020
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	05/09/2020	05/09/2020
					CT233	Surgical	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	05/09/2020	05/09/2020
							F1901	Additional Surgery	A070504	Premature Discharge of Battery	05/09/2020	05/09/2020
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	05/09/2020	05/09/2020
					CT2406	Implantable	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	05/09/2020	05/09/2020
							F1901	Additional Surgery	A070504	Premature Discharge of Battery	05/09/2020	05/09/2020
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	05/09/2020	05/09/2020
					CT2407	Active implantable	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	05/09/2020	05/09/2020
							F1901	Additional Surgery	A070504	Premature Discharge of Battery	05/09/2020	05/09/2020
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	05/09/2020	05/09/2020
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	05/09/2020	05/09/2020
							F1901	Additional Surgery	A070504	Premature Discharge of Battery	05/09/2020	05/09/2020
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	05/09/2020	05/09/2020
					CT2727	Active implant devices/ancillaries	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	05/09/2020	05/09/2020
							F1901	Additional Surgery	A070504	Premature Discharge of Battery	05/09/2020	05/09/2020
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	05/09/2020	05/09/2020
					CT2750	Implantable cardiac electrical stimulators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	05/09/2020	05/09/2020
							F1901	Additional Surgery	A070504	Premature Discharge of Battery	05/09/2020	05/09/2020
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	05/09/2020	05/09/2020
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	05/09/2020	05/09/2020
							F1901	Additional Surgery	A070504	Premature Discharge of Battery	05/09/2020	05/09/2020
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	05/09/2020	05/09/2020
					CT2985	Active	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	05/09/2020	05/09/2020
							F1901	Additional Surgery	A070504	Premature Discharge of Battery	05/09/2020	05/09/2020
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	05/09/2020	05/09/2020
					CT3001	Implantable metal	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	05/09/2020	05/09/2020
							F1901	Additional Surgery	A070504	Premature Discharge of Battery	05/09/2020	05/09/2020

						F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	05/09/2020	05/09/2020
					CT3002	Implantable synthetic polymer	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	05/09/2020	05/09/2020
							F1901	Additional Surgery	A070504	Premature Discharge of Battery	05/09/2020	05/09/2020
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	05/09/2020	05/09/2020
					CT3006	Implantable inorganic material	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	05/09/2020	05/09/2020
							F1901	Additional Surgery	A070504	Premature Discharge of Battery	05/09/2020	05/09/2020
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	05/09/2020	05/09/2020
					CT314	Electrically powered	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	05/09/2020	05/09/2020
							F1901	Additional Surgery	A070504	Premature Discharge of Battery	05/09/2020	05/09/2020
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	05/09/2020	05/09/2020
					CT321	Long-term surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	05/09/2020	05/09/2020
							F1901	Additional Surgery	A070504	Premature Discharge of Battery	05/09/2020	05/09/2020
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	05/09/2020	05/09/2020
					CT449	Generators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	05/09/2020	05/09/2020
							F1901	Additional Surgery	A070504	Premature Discharge of Battery	05/09/2020	05/09/2020
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	05/09/2020	05/09/2020
					CT981	Single use	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	05/09/2020	05/09/2020
							F1901	Additional Surgery	A070504	Premature Discharge of Battery	05/09/2020	05/09/2020
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	05/09/2020	05/09/2020
					CT983	Surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	05/09/2020	05/09/2020
							F1901	Additional Surgery	A070504	Premature Discharge of Battery	05/09/2020	05/09/2020
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	05/09/2020	05/09/2020
11808	2020	Death	35852	Single-chamber implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E060109	Tachycardia	A070910	Undersensing		
							E2401	Insufficient Information	A070910	Undersensing		
							F02	Death	A070910	Undersensing		
11809	2020	All other reportable incidents	47267	Single-chamber implantable pacemaker, rate-responsive	CT141	Electrodes/Leads and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070909	Oversensing		
									A0722	Impedance Problem		
					CT145	Cardiovascular devices	E2403	No Clinical Signs, Symptoms or Conditions	A070909	Oversensing		
									A0722	Impedance Problem		
					CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070909	Oversensing		
									A0722	Impedance Problem		
					CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A070909	Oversensing		
									A0722	Impedance Problem		
					CT233	Surgical	E2403	No Clinical Signs, Symptoms or Conditions	A070909	Oversensing		
									A0722	Impedance Problem		
					CT2406	Implantable	E2403	No Clinical Signs, Symptoms or Conditions	A070909	Oversensing		
									A0722	Impedance Problem		
					CT2407	Active implantable	E2403	No Clinical Signs, Symptoms or Conditions	A070909	Oversensing		
									A0722	Impedance Problem		
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A070909	Oversensing		
									A0722	Impedance Problem		
					CT2727	Active implant devices/ancillaries	E2403	No Clinical Signs, Symptoms or Conditions	A070909	Oversensing		
									A0722	Impedance Problem		
					CT2750	Implantable cardiac electrical stimulators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070909	Oversensing		
									A0722	Impedance Problem		
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A070909	Oversensing		
									A0722	Impedance Problem		
					CT2985	Active	E2403	No Clinical Signs, Symptoms or Conditions	A070909	Oversensing		
									A0722	Impedance Problem		
					CT3001	Implantable metal	E2403	No Clinical Signs, Symptoms or Conditions	A070909	Oversensing		
									A0722	Impedance Problem		
					CT3002	Implantable synthetic polymer	E2403	No Clinical Signs, Symptoms or Conditions	A070909	Oversensing		
									A0722	Impedance Problem		
					CT3006	Implantable inorganic material	E2403	No Clinical Signs, Symptoms or Conditions	A070909	Oversensing		
									A0722	Impedance Problem		
					CT314	Electrically powered	E2403	No Clinical Signs, Symptoms or Conditions	A070909	Oversensing		
									A0722	Impedance Problem		
					CT321	Long-term surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A070909	Oversensing		
									A0722	Impedance Problem		
					CT449	Generators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070909	Oversensing		
									A0722	Impedance Problem		
					CT981	Single use	E2403	No Clinical Signs, Symptoms or Conditions	A070909	Oversensing		
									A0722	Impedance Problem		
					CT983	Surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A070909	Oversensing		
									A0722	Impedance Problem		
11810	2021	All other reportable incidents	35223	Endocardial/interventricular septal pacing lead	CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A040101	Fracture		
									A090801	Signal Artifact/Noise		
							F26	No Health Consequences or Impact	A040101	Fracture		
									A090801	Signal Artifact/Noise		
11811	2019	All other reportable incidents	35223	Endocardial/interventricular septal pacing lead	CT1684	Cardiac pacemakers and associated devices	F1905	Device Revision or Replacement	A072201	High Impedance	20/06/2019	
11812	2021	All other reportable incidents	47263	Cardiac resynchronization therapy implantable pacemaker	CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	18/01/2022	18/01/2022
							F1901	Additional Surgery	A070504	Premature Discharge of Battery	18/01/2022	18/01/2022
					CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	18/01/2022	18/01/2022
							F1901	Additional Surgery	A070504	Premature Discharge of Battery	18/01/2022	18/01/2022
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	18/01/2022	18/01/2022
							F1901	Additional Surgery	A070504	Premature Discharge of Battery	18/01/2022	18/01/2022
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	18/01/2022	18/01/2022
							F1901	Additional Surgery	A070504	Premature Discharge of Battery	18/01/2022	18/01/2022
11813	2019	All other reportable incidents	35223	Endocardial/interventricular septal pacing lead	CT1684	Cardiac pacemakers and associated devices	F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A010402	Migration	20/05/2019	
							F1905	Device Revision or Replacement	A010402	Migration	20/05/2019	
11814	2020	All other reportable incidents	36241	Pacemaker/defibrillator lead	CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A0203	Defective Device	02/06/2020	02/06/2020
							F05	Delay to Treatment/ Therapy	A0203	Defective Device	02/06/2020	02/06/2020
							F26	No Health Consequences or Impact	A0203	Defective Device	02/06/2020	02/06/2020
11815	2019	All other reportable incidents	36241	Pacemaker/defibrillator lead	CT1684	Cardiac pacemakers and associated devices	F1905	Device Revision or Replacement	A071302	Inappropriate/Inadequate Shock/Stimulation	14/11/2019	14/11/2019
11816	2020	All other reportable incidents	47267	Single-chamber implantable pacemaker, rate-responsive	CT141	Electrodes/Leads and associated devices	E011903	Syncope/Fainting	A09	Output Problem		
							E2403	No Clinical Signs, Symptoms or Conditions	A09	Output Problem		
							F1203	Life Threatening Illness or Injury	A09	Output Problem		

					CT145	Cardiovascular devices	E011903	Syncope/Fainting	A09	Output Problem		
							E2403	No Clinical Signs, Symptoms or Conditions	A09	Output Problem		
							F1203	Life Threatening Illness or Injury	A09	Output Problem		
					CT1684	Cardiac pacemakers and associated devices	E011903	Syncope/Fainting	A09	Output Problem		
							E2403	No Clinical Signs, Symptoms or Conditions	A09	Output Problem		
							F1203	Life Threatening Illness or Injury	A09	Output Problem		
					CT1685	Cardiac pacemakers	E011903	Syncope/Fainting	A09	Output Problem		
							E2403	No Clinical Signs, Symptoms or Conditions	A09	Output Problem		
							F1203	Life Threatening Illness or Injury	A09	Output Problem		
					CT233	Surgical	E011903	Syncope/Fainting	A09	Output Problem		
							E2403	No Clinical Signs, Symptoms or Conditions	A09	Output Problem		
							F1203	Life Threatening Illness or Injury	A09	Output Problem		
					CT2406	Implantable	E011903	Syncope/Fainting	A09	Output Problem		
							E2403	No Clinical Signs, Symptoms or Conditions	A09	Output Problem		
							F1203	Life Threatening Illness or Injury	A09	Output Problem		
					CT2407	Active implantable	E011903	Syncope/Fainting	A09	Output Problem		
							E2403	No Clinical Signs, Symptoms or Conditions	A09	Output Problem		
							F1203	Life Threatening Illness or Injury	A09	Output Problem		
					CT257	Implantable cardiac pacemakers	E011903	Syncope/Fainting	A09	Output Problem		
							E2403	No Clinical Signs, Symptoms or Conditions	A09	Output Problem		
							F1203	Life Threatening Illness or Injury	A09	Output Problem		
					CT2727	Active implant devices/ancillaries	E011903	Syncope/Fainting	A09	Output Problem		
							E2403	No Clinical Signs, Symptoms or Conditions	A09	Output Problem		
							F1203	Life Threatening Illness or Injury	A09	Output Problem		
					CT2750	Implantable cardiac electrical stimulators and associated devices	E011903	Syncope/Fainting	A09	Output Problem		
							E2403	No Clinical Signs, Symptoms or Conditions	A09	Output Problem		
							F1203	Life Threatening Illness or Injury	A09	Output Problem		
					CT2866	Implantable cardiac electrical stimulators	E011903	Syncope/Fainting	A09	Output Problem		
							E2403	No Clinical Signs, Symptoms or Conditions	A09	Output Problem		
							F1203	Life Threatening Illness or Injury	A09	Output Problem		
					CT2985	Active	E011903	Syncope/Fainting	A09	Output Problem		
							E2403	No Clinical Signs, Symptoms or Conditions	A09	Output Problem		
							F1203	Life Threatening Illness or Injury	A09	Output Problem		
					CT3001	Implantable metal	E011903	Syncope/Fainting	A09	Output Problem		
							E2403	No Clinical Signs, Symptoms or Conditions	A09	Output Problem		
							F1203	Life Threatening Illness or Injury	A09	Output Problem		
					CT3002	Implantable synthetic polymer	E011903	Syncope/Fainting	A09	Output Problem		
							E2403	No Clinical Signs, Symptoms or Conditions	A09	Output Problem		
							F1203	Life Threatening Illness or Injury	A09	Output Problem		
					CT3006	Implantable inorganic material	E011903	Syncope/Fainting	A09	Output Problem		
							E2403	No Clinical Signs, Symptoms or Conditions	A09	Output Problem		
							F1203	Life Threatening Illness or Injury	A09	Output Problem		
					CT314	Electrically powered	E011903	Syncope/Fainting	A09	Output Problem		
							E2403	No Clinical Signs, Symptoms or Conditions	A09	Output Problem		
							F1203	Life Threatening Illness or Injury	A09	Output Problem		
					CT321	Long-term surgical invasive	E011903	Syncope/Fainting	A09	Output Problem		
							E2403	No Clinical Signs, Symptoms or Conditions	A09	Output Problem		
							F1203	Life Threatening Illness or Injury	A09	Output Problem		
					CT449	Generators and associated devices	E011903	Syncope/Fainting	A09	Output Problem		
							E2403	No Clinical Signs, Symptoms or Conditions	A09	Output Problem		
							F1203	Life Threatening Illness or Injury	A09	Output Problem		
					CT981	Single use	E011903	Syncope/Fainting	A09	Output Problem		
							E2403	No Clinical Signs, Symptoms or Conditions	A09	Output Problem		
							F1203	Life Threatening Illness or Injury	A09	Output Problem		
					CT983	Surgical invasive	E011903	Syncope/Fainting	A09	Output Problem		
							E2403	No Clinical Signs, Symptoms or Conditions	A09	Output Problem		
							F1203	Life Threatening Illness or Injury	A09	Output Problem		
11817	2019	All other reportable incidents	47265	Dual-chamber implantable pacemaker, rate-responsive	CT141	Electrodes/Leads and associated devices	F1905	Device Revision or Replacement	A070909	Oversensing	15/10/2019	15/10/2019
					CT145	Cardiovascular devices	F1905	Device Revision or Replacement	A070909	Oversensing	15/10/2019	15/10/2019
					CT1684	Cardiac pacemakers and associated devices	F1905	Device Revision or Replacement	A070909	Oversensing	15/10/2019	15/10/2019
					CT1685	Cardiac pacemakers	F1905	Device Revision or Replacement	A070909	Oversensing	15/10/2019	15/10/2019
					CT233	Surgical	F1905	Device Revision or Replacement	A070909	Oversensing	15/10/2019	15/10/2019
					CT2406	Implantable	F1905	Device Revision or Replacement	A070909	Oversensing	15/10/2019	15/10/2019
					CT2407	Active implantable	F1905	Device Revision or Replacement	A070909	Oversensing	15/10/2019	15/10/2019
					CT257	Implantable cardiac pacemakers	F1905	Device Revision or Replacement	A070909	Oversensing	15/10/2019	15/10/2019
					CT2727	Active implant devices/ancillaries	F1905	Device Revision or Replacement	A070909	Oversensing	15/10/2019	15/10/2019
					CT2750	Implantable cardiac electrical stimulators and associated devices	F1905	Device Revision or Replacement	A070909	Oversensing	15/10/2019	15/10/2019
					CT2866	Implantable cardiac electrical stimulators	F1905	Device Revision or Replacement	A070909	Oversensing	15/10/2019	15/10/2019
					CT2985	Active	F1905	Device Revision or Replacement	A070909	Oversensing	15/10/2019	15/10/2019
					CT3001	Implantable metal	F1905	Device Revision or Replacement	A070909	Oversensing	15/10/2019	15/10/2019
					CT3002	Implantable synthetic polymer	F1905	Device Revision or Replacement	A070909	Oversensing	15/10/2019	15/10/2019
					CT3006	Implantable inorganic material	F1905	Device Revision or Replacement	A070909	Oversensing	15/10/2019	15/10/2019
					CT314	Electrically powered	F1905	Device Revision or Replacement	A070909	Oversensing	15/10/2019	15/10/2019
					CT321	Long-term surgical invasive	F1905	Device Revision or Replacement	A070909	Oversensing	15/10/2019	15/10/2019
					CT449	Generators and associated devices	F1905	Device Revision or Replacement	A070909	Oversensing	15/10/2019	15/10/2019
					CT981	Single use	F1905	Device Revision or Replacement	A070909	Oversensing	15/10/2019	15/10/2019
					CT983	Surgical invasive	F1905	Device Revision or Replacement	A070909	Oversensing	15/10/2019	15/10/2019
11818	2019	All other reportable incidents	47265	Dual-chamber implantable pacemaker, rate-responsive	CT141	Electrodes/Leads and associated devices	F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	18/11/2019	
					CT145	Cardiovascular devices	F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	18/11/2019	
					CT1684	Cardiac pacemakers and associated devices	F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	18/11/2019	
					CT1685	Cardiac pacemakers	F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	18/11/2019	
					CT233	Surgical	F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	18/11/2019	
					CT2406	Implantable	F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	18/11/2019	
					CT2407	Active implantable	F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	18/11/2019	
					CT257	Implantable cardiac pacemakers	F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	18/11/2019	
					CT2727	Active implant devices/ancillaries	F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	18/11/2019	

					CT2750	Implantable cardiac electrical stimulators and associated devices	F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	18/11/2019		
					CT2866	Implantable cardiac electrical stimulators	F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	18/11/2019		
					CT2985	Active	F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	18/11/2019		
					CT3001	Implantable metal	F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	18/11/2019		
					CT3002	Implantable synthetic polymer	F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	18/11/2019		
					CT3006	Implantable inorganic material	F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	18/11/2019		
					CT314	Electrically powered	F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	18/11/2019		
					CT321	Long-term surgical invasive	F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	18/11/2019		
					CT449	Generators and associated devices	F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	18/11/2019		
					CT981	Single use	F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	18/11/2019		
					CT983	Surgical invasive	F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	18/11/2019		
11819	2020	All other reportable incidents	35852	Single-chamber implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A070909	Oversensing			
							F26	No Health Consequences or Impact	A070909	Oversensing			
11820	2020	All other reportable incidents	35852	Single-chamber implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E2104	Electric Shock	A0712	Pacing Problem			
									A071302	Inappropriate/Inadequate Shock/Stimulation			
							F26	No Health Consequences or Impact	A0712	Pacing Problem			
									A071302	Inappropriate/Inadequate Shock/Stimulation			
11821	2019	All other reportable incidents	47265	Dual-chamber implantable pacemaker, rate-responsive	CT141	Electrodes/Leads and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem	30/03/2021	30/03/2021	
									A070504	Premature Discharge of Battery	30/03/2021	30/03/2021	
							F08	Hospitalization or Prolonged Hospitalization	A0705	Battery Problem	30/03/2021	30/03/2021	
									A070504	Premature Discharge of Battery	30/03/2021	30/03/2021	
							F12	Serious Injury/ Illness/ Impairment	A0705	Battery Problem	30/03/2021	30/03/2021	
									A070504	Premature Discharge of Battery	30/03/2021	30/03/2021	
							F1903	Device Explantation	A0705	Battery Problem	30/03/2021	30/03/2021	
									A070504	Premature Discharge of Battery	30/03/2021	30/03/2021	
					CT145	Cardiovascular devices	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem	30/03/2021	30/03/2021	
									A070504	Premature Discharge of Battery	30/03/2021	30/03/2021	
							F08	Hospitalization or Prolonged Hospitalization	A0705	Battery Problem	30/03/2021	30/03/2021	
									A070504	Premature Discharge of Battery	30/03/2021	30/03/2021	
							F12	Serious Injury/ Illness/ Impairment	A0705	Battery Problem	30/03/2021	30/03/2021	
									A070504	Premature Discharge of Battery	30/03/2021	30/03/2021	
							F1903	Device Explantation	A0705	Battery Problem	30/03/2021	30/03/2021	
									A070504	Premature Discharge of Battery	30/03/2021	30/03/2021	
					CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem	30/03/2021	30/03/2021	
									A070504	Premature Discharge of Battery	30/03/2021	30/03/2021	
							F08	Hospitalization or Prolonged Hospitalization	A0705	Battery Problem	30/03/2021	30/03/2021	
									A070504	Premature Discharge of Battery	30/03/2021	30/03/2021	
							F12	Serious Injury/ Illness/ Impairment	A0705	Battery Problem	30/03/2021	30/03/2021	
									A070504	Premature Discharge of Battery	30/03/2021	30/03/2021	
							F1903	Device Explantation	A0705	Battery Problem	30/03/2021	30/03/2021	
									A070504	Premature Discharge of Battery	30/03/2021	30/03/2021	
					CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem	30/03/2021	30/03/2021	
									A070504	Premature Discharge of Battery	30/03/2021	30/03/2021	
							F08	Hospitalization or Prolonged Hospitalization	A0705	Battery Problem	30/03/2021	30/03/2021	
									A070504	Premature Discharge of Battery	30/03/2021	30/03/2021	
							F12	Serious Injury/ Illness/ Impairment	A0705	Battery Problem	30/03/2021	30/03/2021	
									A070504	Premature Discharge of Battery	30/03/2021	30/03/2021	
							F1903	Device Explantation	A0705	Battery Problem	30/03/2021	30/03/2021	
									A070504	Premature Discharge of Battery	30/03/2021	30/03/2021	
					CT233	Surgical	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem	30/03/2021	30/03/2021	
									A070504	Premature Discharge of Battery	30/03/2021	30/03/2021	
							F08	Hospitalization or Prolonged Hospitalization	A0705	Battery Problem	30/03/2021	30/03/2021	
									A070504	Premature Discharge of Battery	30/03/2021	30/03/2021	
							F12	Serious Injury/ Illness/ Impairment	A0705	Battery Problem	30/03/2021	30/03/2021	
									A070504	Premature Discharge of Battery	30/03/2021	30/03/2021	
							F1903	Device Explantation	A0705	Battery Problem	30/03/2021	30/03/2021	
									A070504	Premature Discharge of Battery	30/03/2021	30/03/2021	
					CT2406	Implantable	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem	30/03/2021	30/03/2021	
									A070504	Premature Discharge of Battery	30/03/2021	30/03/2021	
							F08	Hospitalization or Prolonged Hospitalization	A0705	Battery Problem	30/03/2021	30/03/2021	
									A070504	Premature Discharge of Battery	30/03/2021	30/03/2021	
							F12	Serious Injury/ Illness/ Impairment	A0705	Battery Problem	30/03/2021	30/03/2021	
									A070504	Premature Discharge of Battery	30/03/2021	30/03/2021	
							F1903	Device Explantation	A0705	Battery Problem	30/03/2021	30/03/2021	
									A070504	Premature Discharge of Battery	30/03/2021	30/03/2021	
					CT2407	Active implantable	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem	30/03/2021	30/03/2021	
									A070504	Premature Discharge of Battery	30/03/2021	30/03/2021	
							F08	Hospitalization or Prolonged Hospitalization	A0705	Battery Problem	30/03/2021	30/03/2021	
									A070504	Premature Discharge of Battery	30/03/2021	30/03/2021	
							F12	Serious Injury/ Illness/ Impairment	A0705	Battery Problem	30/03/2021	30/03/2021	
									A070504	Premature Discharge of Battery	30/03/2021	30/03/2021	
							F1903	Device Explantation	A0705	Battery Problem	30/03/2021	30/03/2021	
									A070504	Premature Discharge of Battery	30/03/2021	30/03/2021	
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem	30/03/2021	30/03/2021	
									A070504	Premature Discharge of Battery	30/03/2021	30/03/2021	
							F08	Hospitalization or Prolonged Hospitalization	A0705	Battery Problem	30/03/2021	30/03/2021	
									A070504	Premature Discharge of Battery	30/03/2021	30/03/2021	
							F12	Serious Injury/ Illness/ Impairment	A0705	Battery Problem	30/03/2021	30/03/2021	
									A070504	Premature Discharge of Battery	30/03/2021	30/03/2021	
							F1903	Device Explantation	A0705	Battery Problem	30/03/2021	30/03/2021	
									A070504	Premature Discharge of Battery	30/03/2021	30/03/2021	
					CT2727	Active implant devices/ancillaries	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem	30/03/2021	30/03/2021	
									A070504	Premature Discharge of Battery	30/03/2021	30/03/2021	
							F08	Hospitalization or Prolonged Hospitalization	A0705	Battery Problem	30/03/2021	30/03/2021	
									A070504	Premature Discharge of Battery	30/03/2021	30/03/2021	
							F12	Serious Injury/ Illness/ Impairment	A0705	Battery Problem	30/03/2021	30/03/2021	
									A070504	Premature Discharge of Battery	30/03/2021	30/03/2021	

							F1903	Device Explantation	A0705	Battery Problem	30/03/2021	30/03/2021
									A070504	Premature Discharge of Battery	30/03/2021	30/03/2021
					CT2750	Implantable cardiac electrical stimulators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem	30/03/2021	30/03/2021
									A070504	Premature Discharge of Battery	30/03/2021	30/03/2021
							F08	Hospitalization or Prolonged Hospitalization	A0705	Battery Problem	30/03/2021	30/03/2021
									A070504	Premature Discharge of Battery	30/03/2021	30/03/2021
							F12	Serious Injury/ Illness/ Impairment	A0705	Battery Problem	30/03/2021	30/03/2021
									A070504	Premature Discharge of Battery	30/03/2021	30/03/2021
							F1903	Device Explantation	A0705	Battery Problem	30/03/2021	30/03/2021
									A070504	Premature Discharge of Battery	30/03/2021	30/03/2021
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem	30/03/2021	30/03/2021
									A070504	Premature Discharge of Battery	30/03/2021	30/03/2021
							F08	Hospitalization or Prolonged Hospitalization	A0705	Battery Problem	30/03/2021	30/03/2021
									A070504	Premature Discharge of Battery	30/03/2021	30/03/2021
							F12	Serious Injury/ Illness/ Impairment	A0705	Battery Problem	30/03/2021	30/03/2021
									A070504	Premature Discharge of Battery	30/03/2021	30/03/2021
							F1903	Device Explantation	A0705	Battery Problem	30/03/2021	30/03/2021
									A070504	Premature Discharge of Battery	30/03/2021	30/03/2021
					CT2985	Active	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem	30/03/2021	30/03/2021
									A070504	Premature Discharge of Battery	30/03/2021	30/03/2021
							F08	Hospitalization or Prolonged Hospitalization	A0705	Battery Problem	30/03/2021	30/03/2021
									A070504	Premature Discharge of Battery	30/03/2021	30/03/2021
							F12	Serious Injury/ Illness/ Impairment	A0705	Battery Problem	30/03/2021	30/03/2021
									A070504	Premature Discharge of Battery	30/03/2021	30/03/2021
							F1903	Device Explantation	A0705	Battery Problem	30/03/2021	30/03/2021
									A070504	Premature Discharge of Battery	30/03/2021	30/03/2021
					CT3001	Implantable metal	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem	30/03/2021	30/03/2021
									A070504	Premature Discharge of Battery	30/03/2021	30/03/2021
							F08	Hospitalization or Prolonged Hospitalization	A0705	Battery Problem	30/03/2021	30/03/2021
									A070504	Premature Discharge of Battery	30/03/2021	30/03/2021
							F12	Serious Injury/ Illness/ Impairment	A0705	Battery Problem	30/03/2021	30/03/2021
									A070504	Premature Discharge of Battery	30/03/2021	30/03/2021
							F1903	Device Explantation	A0705	Battery Problem	30/03/2021	30/03/2021
									A070504	Premature Discharge of Battery	30/03/2021	30/03/2021
					CT3002	Implantable synthetic polymer	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem	30/03/2021	30/03/2021
									A070504	Premature Discharge of Battery	30/03/2021	30/03/2021
							F08	Hospitalization or Prolonged Hospitalization	A0705	Battery Problem	30/03/2021	30/03/2021
									A070504	Premature Discharge of Battery	30/03/2021	30/03/2021
							F12	Serious Injury/ Illness/ Impairment	A0705	Battery Problem	30/03/2021	30/03/2021
									A070504	Premature Discharge of Battery	30/03/2021	30/03/2021
							F1903	Device Explantation	A0705	Battery Problem	30/03/2021	30/03/2021
									A070504	Premature Discharge of Battery	30/03/2021	30/03/2021
					CT3006	Implantable inorganic material	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem	30/03/2021	30/03/2021
									A070504	Premature Discharge of Battery	30/03/2021	30/03/2021
							F08	Hospitalization or Prolonged Hospitalization	A0705	Battery Problem	30/03/2021	30/03/2021
									A070504	Premature Discharge of Battery	30/03/2021	30/03/2021
							F12	Serious Injury/ Illness/ Impairment	A0705	Battery Problem	30/03/2021	30/03/2021
									A070504	Premature Discharge of Battery	30/03/2021	30/03/2021
							F1903	Device Explantation	A0705	Battery Problem	30/03/2021	30/03/2021
									A070504	Premature Discharge of Battery	30/03/2021	30/03/2021
					CT314	Electrically powered	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem	30/03/2021	30/03/2021
									A070504	Premature Discharge of Battery	30/03/2021	30/03/2021
							F08	Hospitalization or Prolonged Hospitalization	A0705	Battery Problem	30/03/2021	30/03/2021
									A070504	Premature Discharge of Battery	30/03/2021	30/03/2021
							F12	Serious Injury/ Illness/ Impairment	A0705	Battery Problem	30/03/2021	30/03/2021
									A070504	Premature Discharge of Battery	30/03/2021	30/03/2021
							F1903	Device Explantation	A0705	Battery Problem	30/03/2021	30/03/2021
									A070504	Premature Discharge of Battery	30/03/2021	30/03/2021
					CT321	Long-term surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem	30/03/2021	30/03/2021
									A070504	Premature Discharge of Battery	30/03/2021	30/03/2021
							F08	Hospitalization or Prolonged Hospitalization	A0705	Battery Problem	30/03/2021	30/03/2021
									A070504	Premature Discharge of Battery	30/03/2021	30/03/2021
							F12	Serious Injury/ Illness/ Impairment	A0705	Battery Problem	30/03/2021	30/03/2021
									A070504	Premature Discharge of Battery	30/03/2021	30/03/2021
							F1903	Device Explantation	A0705	Battery Problem	30/03/2021	30/03/2021
									A070504	Premature Discharge of Battery	30/03/2021	30/03/2021
					CT449	Generators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem	30/03/2021	30/03/2021
									A070504	Premature Discharge of Battery	30/03/2021	30/03/2021
							F08	Hospitalization or Prolonged Hospitalization	A0705	Battery Problem	30/03/2021	30/03/2021
									A070504	Premature Discharge of Battery	30/03/2021	30/03/2021
							F12	Serious Injury/ Illness/ Impairment	A0705	Battery Problem	30/03/2021	30/03/2021
									A070504	Premature Discharge of Battery	30/03/2021	30/03/2021
							F1903	Device Explantation	A0705	Battery Problem	30/03/2021	30/03/2021
									A070504	Premature Discharge of Battery	30/03/2021	30/03/2021
					CT981	Single use	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem	30/03/2021	30/03/2021
									A070504	Premature Discharge of Battery	30/03/2021	30/03/2021
							F08	Hospitalization or Prolonged Hospitalization	A0705	Battery Problem	30/03/2021	30/03/2021
									A070504	Premature Discharge of Battery	30/03/2021	30/03/2021
							F12	Serious Injury/ Illness/ Impairment	A0705	Battery Problem	30/03/2021	30/03/2021
									A070504	Premature Discharge of Battery	30/03/2021	30/03/2021
							F1903	Device Explantation	A0705	Battery Problem	30/03/2021	30/03/2021
									A070504	Premature Discharge of Battery	30/03/2021	30/03/2021
					CT983	Surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem	30/03/2021	30/03/2021
									A070504	Premature Discharge of Battery	30/03/2021	30/03/2021
							F08	Hospitalization or Prolonged Hospitalization	A0705	Battery Problem	30/03/2021	30/03/2021
									A070504	Premature Discharge of Battery	30/03/2021	30/03/2021
							F12	Serious Injury/ Illness/ Impairment	A0705	Battery Problem	30/03/2021	30/03/2021
									A070504	Premature Discharge of Battery	30/03/2021	30/03/2021

							F1903	Device Explantation	A0705	Battery Problem	30/03/2021	30/03/2021
									A070504	Premature Discharge of Battery	30/03/2021	30/03/2021
11822	2020	All other reportable incidents	47265	Dual-chamber implantable pacemaker, rate-responsive	CT141	Electrodes/Leads and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem	17/06/2020	17/06/2020
									A071102	Failure to Interrogate	17/06/2020	17/06/2020
									A071204	No Pacing	17/06/2020	17/06/2020
							F1901	Additional Surgery	A0705	Battery Problem	17/06/2020	17/06/2020
									A071102	Failure to Interrogate	17/06/2020	17/06/2020
									A071204	No Pacing	17/06/2020	17/06/2020
							F1905	Device Revision or Replacement	A0705	Battery Problem	17/06/2020	17/06/2020
									A071102	Failure to Interrogate	17/06/2020	17/06/2020
									A071204	No Pacing	17/06/2020	17/06/2020
					CT145	Cardiovascular devices	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem	17/06/2020	17/06/2020
									A071102	Failure to Interrogate	17/06/2020	17/06/2020
									A071204	No Pacing	17/06/2020	17/06/2020
							F1901	Additional Surgery	A0705	Battery Problem	17/06/2020	17/06/2020
									A071102	Failure to Interrogate	17/06/2020	17/06/2020
									A071204	No Pacing	17/06/2020	17/06/2020
							F1905	Device Revision or Replacement	A0705	Battery Problem	17/06/2020	17/06/2020
									A071102	Failure to Interrogate	17/06/2020	17/06/2020
									A071204	No Pacing	17/06/2020	17/06/2020
					CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem	17/06/2020	17/06/2020
									A071102	Failure to Interrogate	17/06/2020	17/06/2020
									A071204	No Pacing	17/06/2020	17/06/2020
							F1901	Additional Surgery	A0705	Battery Problem	17/06/2020	17/06/2020
									A071102	Failure to Interrogate	17/06/2020	17/06/2020
									A071204	No Pacing	17/06/2020	17/06/2020
							F1905	Device Revision or Replacement	A0705	Battery Problem	17/06/2020	17/06/2020
									A071102	Failure to Interrogate	17/06/2020	17/06/2020
									A071204	No Pacing	17/06/2020	17/06/2020
					CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem	17/06/2020	17/06/2020
									A071102	Failure to Interrogate	17/06/2020	17/06/2020
									A071204	No Pacing	17/06/2020	17/06/2020
							F1901	Additional Surgery	A0705	Battery Problem	17/06/2020	17/06/2020
									A071102	Failure to Interrogate	17/06/2020	17/06/2020
									A071204	No Pacing	17/06/2020	17/06/2020
							F1905	Device Revision or Replacement	A0705	Battery Problem	17/06/2020	17/06/2020
									A071102	Failure to Interrogate	17/06/2020	17/06/2020
									A071204	No Pacing	17/06/2020	17/06/2020
					CT233	Surgical	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem	17/06/2020	17/06/2020
									A071102	Failure to Interrogate	17/06/2020	17/06/2020
									A071204	No Pacing	17/06/2020	17/06/2020
							F1901	Additional Surgery	A0705	Battery Problem	17/06/2020	17/06/2020
									A071102	Failure to Interrogate	17/06/2020	17/06/2020
									A071204	No Pacing	17/06/2020	17/06/2020
							F1905	Device Revision or Replacement	A0705	Battery Problem	17/06/2020	17/06/2020
									A071102	Failure to Interrogate	17/06/2020	17/06/2020
									A071204	No Pacing	17/06/2020	17/06/2020
					CT2406	Implantable	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem	17/06/2020	17/06/2020
									A071102	Failure to Interrogate	17/06/2020	17/06/2020
									A071204	No Pacing	17/06/2020	17/06/2020
							F1901	Additional Surgery	A0705	Battery Problem	17/06/2020	17/06/2020
									A071102	Failure to Interrogate	17/06/2020	17/06/2020
									A071204	No Pacing	17/06/2020	17/06/2020
							F1905	Device Revision or Replacement	A0705	Battery Problem	17/06/2020	17/06/2020
									A071102	Failure to Interrogate	17/06/2020	17/06/2020
									A071204	No Pacing	17/06/2020	17/06/2020
					CT2407	Active implantable	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem	17/06/2020	17/06/2020
									A071102	Failure to Interrogate	17/06/2020	17/06/2020
									A071204	No Pacing	17/06/2020	17/06/2020
							F1901	Additional Surgery	A0705	Battery Problem	17/06/2020	17/06/2020
									A071102	Failure to Interrogate	17/06/2020	17/06/2020
									A071204	No Pacing	17/06/2020	17/06/2020
							F1905	Device Revision or Replacement	A0705	Battery Problem	17/06/2020	17/06/2020
									A071102	Failure to Interrogate	17/06/2020	17/06/2020
									A071204	No Pacing	17/06/2020	17/06/2020
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem	17/06/2020	17/06/2020
									A071102	Failure to Interrogate	17/06/2020	17/06/2020
									A071204	No Pacing	17/06/2020	17/06/2020
							F1901	Additional Surgery	A0705	Battery Problem	17/06/2020	17/06/2020
									A071102	Failure to Interrogate	17/06/2020	17/06/2020
									A071204	No Pacing	17/06/2020	17/06/2020
							F1905	Device Revision or Replacement	A0705	Battery Problem	17/06/2020	17/06/2020
									A071102	Failure to Interrogate	17/06/2020	17/06/2020
									A071204	No Pacing	17/06/2020	17/06/2020
					CT2727	Active implant devices/ancillaries	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem	17/06/2020	17/06/2020
									A071102	Failure to Interrogate	17/06/2020	17/06/2020
									A071204	No Pacing	17/06/2020	17/06/2020
							F1901	Additional Surgery	A0705	Battery Problem	17/06/2020	17/06/2020
									A071102	Failure to Interrogate	17/06/2020	17/06/2020
									A071204	No Pacing	17/06/2020	17/06/2020
							F1905	Device Revision or Replacement	A0705	Battery Problem	17/06/2020	17/06/2020
									A071102	Failure to Interrogate	17/06/2020	17/06/2020
									A071204	No Pacing	17/06/2020	17/06/2020
					CT2750	Implantable cardiac electrical stimulators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem	17/06/2020	17/06/2020
									A071102	Failure to Interrogate	17/06/2020	17/06/2020
									A071204	No Pacing	17/06/2020	17/06/2020
							F1901	Additional Surgery	A0705	Battery Problem	17/06/2020	17/06/2020
									A071102	Failure to Interrogate	17/06/2020	17/06/2020

								A071204	No Pacing	17/06/2020	17/06/2020
						F1905	Device Revision or Replacement	A0705	Battery Problem	17/06/2020	17/06/2020
								A071102	Failure to Interrogate	17/06/2020	17/06/2020
								A071204	No Pacing	17/06/2020	17/06/2020
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem	17/06/2020	17/06/2020
									A071102	Failure to Interrogate	17/06/2020	17/06/2020
									A071204	No Pacing	17/06/2020	17/06/2020
							F1901	Additional Surgery	A0705	Battery Problem	17/06/2020	17/06/2020
									A071102	Failure to Interrogate	17/06/2020	17/06/2020
									A071204	No Pacing	17/06/2020	17/06/2020
							F1905	Device Revision or Replacement	A0705	Battery Problem	17/06/2020	17/06/2020
									A071102	Failure to Interrogate	17/06/2020	17/06/2020
									A071204	No Pacing	17/06/2020	17/06/2020
					CT2985	Active	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem	17/06/2020	17/06/2020
									A071102	Failure to Interrogate	17/06/2020	17/06/2020
									A071204	No Pacing	17/06/2020	17/06/2020
							F1901	Additional Surgery	A0705	Battery Problem	17/06/2020	17/06/2020
									A071102	Failure to Interrogate	17/06/2020	17/06/2020
									A071204	No Pacing	17/06/2020	17/06/2020
							F1905	Device Revision or Replacement	A0705	Battery Problem	17/06/2020	17/06/2020
									A071102	Failure to Interrogate	17/06/2020	17/06/2020
									A071204	No Pacing	17/06/2020	17/06/2020
					CT3001	Implantable metal	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem	17/06/2020	17/06/2020
									A071102	Failure to Interrogate	17/06/2020	17/06/2020
									A071204	No Pacing	17/06/2020	17/06/2020
							F1901	Additional Surgery	A0705	Battery Problem	17/06/2020	17/06/2020
									A071102	Failure to Interrogate	17/06/2020	17/06/2020
									A071204	No Pacing	17/06/2020	17/06/2020
							F1905	Device Revision or Replacement	A0705	Battery Problem	17/06/2020	17/06/2020
									A071102	Failure to Interrogate	17/06/2020	17/06/2020
									A071204	No Pacing	17/06/2020	17/06/2020
					CT3002	Implantable synthetic polymer	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem	17/06/2020	17/06/2020
									A071102	Failure to Interrogate	17/06/2020	17/06/2020
									A071204	No Pacing	17/06/2020	17/06/2020
							F1901	Additional Surgery	A0705	Battery Problem	17/06/2020	17/06/2020
									A071102	Failure to Interrogate	17/06/2020	17/06/2020
									A071204	No Pacing	17/06/2020	17/06/2020
							F1905	Device Revision or Replacement	A0705	Battery Problem	17/06/2020	17/06/2020
									A071102	Failure to Interrogate	17/06/2020	17/06/2020
									A071204	No Pacing	17/06/2020	17/06/2020
					CT3006	Implantable inorganic material	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem	17/06/2020	17/06/2020
									A071102	Failure to Interrogate	17/06/2020	17/06/2020
									A071204	No Pacing	17/06/2020	17/06/2020
							F1901	Additional Surgery	A0705	Battery Problem	17/06/2020	17/06/2020
									A071102	Failure to Interrogate	17/06/2020	17/06/2020
									A071204	No Pacing	17/06/2020	17/06/2020
							F1905	Device Revision or Replacement	A0705	Battery Problem	17/06/2020	17/06/2020
									A071102	Failure to Interrogate	17/06/2020	17/06/2020
									A071204	No Pacing	17/06/2020	17/06/2020
					CT314	Electrically powered	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem	17/06/2020	17/06/2020
									A071102	Failure to Interrogate	17/06/2020	17/06/2020
									A071204	No Pacing	17/06/2020	17/06/2020
							F1901	Additional Surgery	A0705	Battery Problem	17/06/2020	17/06/2020
									A071102	Failure to Interrogate	17/06/2020	17/06/2020
									A071204	No Pacing	17/06/2020	17/06/2020
							F1905	Device Revision or Replacement	A0705	Battery Problem	17/06/2020	17/06/2020
									A071102	Failure to Interrogate	17/06/2020	17/06/2020
									A071204	No Pacing	17/06/2020	17/06/2020
					CT321	Long-term surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem	17/06/2020	17/06/2020
									A071102	Failure to Interrogate	17/06/2020	17/06/2020
									A071204	No Pacing	17/06/2020	17/06/2020
							F1901	Additional Surgery	A0705	Battery Problem	17/06/2020	17/06/2020
									A071102	Failure to Interrogate	17/06/2020	17/06/2020
									A071204	No Pacing	17/06/2020	17/06/2020
							F1905	Device Revision or Replacement	A0705	Battery Problem	17/06/2020	17/06/2020
									A071102	Failure to Interrogate	17/06/2020	17/06/2020
									A071204	No Pacing	17/06/2020	17/06/2020
					CT449	Generators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem	17/06/2020	17/06/2020
									A071102	Failure to Interrogate	17/06/2020	17/06/2020
									A071204	No Pacing	17/06/2020	17/06/2020
							F1901	Additional Surgery	A0705	Battery Problem	17/06/2020	17/06/2020
									A071102	Failure to Interrogate	17/06/2020	17/06/2020
									A071204	No Pacing	17/06/2020	17/06/2020
							F1905	Device Revision or Replacement	A0705	Battery Problem	17/06/2020	17/06/2020
									A071102	Failure to Interrogate	17/06/2020	17/06/2020
									A071204	No Pacing	17/06/2020	17/06/2020
					CT981	Single use	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem	17/06/2020	17/06/2020
									A071102	Failure to Interrogate	17/06/2020	17/06/2020
									A071204	No Pacing	17/06/2020	17/06/2020
							F1901	Additional Surgery	A0705	Battery Problem	17/06/2020	17/06/2020
									A071102	Failure to Interrogate	17/06/2020	17/06/2020
									A071204	No Pacing	17/06/2020	17/06/2020
							F1905	Device Revision or Replacement	A0705	Battery Problem	17/06/2020	17/06/2020
									A071102	Failure to Interrogate	17/06/2020	17/06/2020
									A071204	No Pacing	17/06/2020	17/06/2020
					CT983	Surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem	17/06/2020	17/06/2020
									A071102	Failure to Interrogate	17/06/2020	17/06/2020
									A071204	No Pacing	17/06/2020	17/06/2020

							F1901	Additional Surgery	A0705	Battery Problem	17/06/2020	17/06/2020
									A071102	Failure to Interrogate	17/06/2020	17/06/2020
									A071204	No Pacing	17/06/2020	17/06/2020
							F1905	Device Revision or Replacement	A0705	Battery Problem	17/06/2020	17/06/2020
									A071102	Failure to Interrogate	17/06/2020	17/06/2020
									A071204	No Pacing	17/06/2020	17/06/2020
11823	2020	All other reportable incidents	36241	Pacemaker/defibrillator lead	CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070909	Oversensing		
									A0712	Pacing Problem		
									A090801	Signal Artifact/Noise		
							F26	No Health Consequences or Impact	A070909	Oversensing		
									A0712	Pacing Problem		
									A090801	Signal Artifact/Noise		
11824	2019	All other reportable incidents	35852	Single-chamber implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A070909	Oversensing		
11825	2022	All other reportable incidents	47265	Dual-chamber implantable pacemaker, rate-responsive	CT141	Electrodes/Leads and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070101	Failure to Capture		
									A070909	Oversensing		
									A071206	Pacing Asynchronously		
							F26	No Health Consequences or Impact	A070101	Failure to Capture		
									A070909	Oversensing		
									A071206	Pacing Asynchronously		
					CT145	Cardiovascular devices	E2403	No Clinical Signs, Symptoms or Conditions	A070101	Failure to Capture		
									A070909	Oversensing		
									A071206	Pacing Asynchronously		
							F26	No Health Consequences or Impact	A070101	Failure to Capture		
									A070909	Oversensing		
									A071206	Pacing Asynchronously		
					CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070101	Failure to Capture		
									A070909	Oversensing		
									A071206	Pacing Asynchronously		
							F26	No Health Consequences or Impact	A070101	Failure to Capture		
									A070909	Oversensing		
									A071206	Pacing Asynchronously		
					CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A070101	Failure to Capture		
									A070909	Oversensing		
									A071206	Pacing Asynchronously		
							F26	No Health Consequences or Impact	A070101	Failure to Capture		
									A070909	Oversensing		
									A071206	Pacing Asynchronously		
					CT233	Surgical	E2403	No Clinical Signs, Symptoms or Conditions	A070101	Failure to Capture		
									A070909	Oversensing		
									A071206	Pacing Asynchronously		
							F26	No Health Consequences or Impact	A070101	Failure to Capture		
									A070909	Oversensing		
									A071206	Pacing Asynchronously		
					CT2406	Implantable	E2403	No Clinical Signs, Symptoms or Conditions	A070101	Failure to Capture		
									A070909	Oversensing		
									A071206	Pacing Asynchronously		
							F26	No Health Consequences or Impact	A070101	Failure to Capture		
									A070909	Oversensing		
									A071206	Pacing Asynchronously		
					CT2407	Active implantable	E2403	No Clinical Signs, Symptoms or Conditions	A070101	Failure to Capture		
									A070909	Oversensing		
									A071206	Pacing Asynchronously		
							F26	No Health Consequences or Impact	A070101	Failure to Capture		
									A070909	Oversensing		
									A071206	Pacing Asynchronously		
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A070101	Failure to Capture		
									A070909	Oversensing		
									A071206	Pacing Asynchronously		
							F26	No Health Consequences or Impact	A070101	Failure to Capture		
									A070909	Oversensing		
									A071206	Pacing Asynchronously		
					CT2727	Active implant devices/ancillaries	E2403	No Clinical Signs, Symptoms or Conditions	A070101	Failure to Capture		
									A070909	Oversensing		
									A071206	Pacing Asynchronously		
							F26	No Health Consequences or Impact	A070101	Failure to Capture		
									A070909	Oversensing		
									A071206	Pacing Asynchronously		
					CT2750	Implantable cardiac electrical stimulators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070101	Failure to Capture		
									A070909	Oversensing		
									A071206	Pacing Asynchronously		
							F26	No Health Consequences or Impact	A070101	Failure to Capture		
									A070909	Oversensing		
									A071206	Pacing Asynchronously		
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A070101	Failure to Capture		
									A070909	Oversensing		
									A071206	Pacing Asynchronously		
							F26	No Health Consequences or Impact	A070101	Failure to Capture		
									A070909	Oversensing		
									A071206	Pacing Asynchronously		
					CT2985	Active	E2403	No Clinical Signs, Symptoms or Conditions	A070101	Failure to Capture		
									A070909	Oversensing		
									A071206	Pacing Asynchronously		
							F26	No Health Consequences or Impact	A070101	Failure to Capture		
									A070909	Oversensing		
									A071206	Pacing Asynchronously		
					CT3001	Implantable metal	E2403	No Clinical Signs, Symptoms or Conditions	A070101	Failure to Capture		
									A070909	Oversensing		
									A071206	Pacing Asynchronously		

							F26	No Health Consequences or Impact	A070101	Failure to Capture			
									A070909	Oversensing			
									A071206	Pacing Asynchronously			
					CT3002	Implantable synthetic polymer	E2403	No Clinical Signs, Symptoms or Conditions	A070101	Failure to Capture			
									A070909	Oversensing			
									A071206	Pacing Asynchronously			
							F26	No Health Consequences or Impact	A070101	Failure to Capture			
									A070909	Oversensing			
									A071206	Pacing Asynchronously			
					CT3006	Implantable inorganic material	E2403	No Clinical Signs, Symptoms or Conditions	A070101	Failure to Capture			
									A070909	Oversensing			
									A071206	Pacing Asynchronously			
							F26	No Health Consequences or Impact	A070101	Failure to Capture			
									A070909	Oversensing			
									A071206	Pacing Asynchronously			
					CT314	Electrically powered	E2403	No Clinical Signs, Symptoms or Conditions	A070101	Failure to Capture			
									A070909	Oversensing			
									A071206	Pacing Asynchronously			
							F26	No Health Consequences or Impact	A070101	Failure to Capture			
									A070909	Oversensing			
									A071206	Pacing Asynchronously			
					CT321	Long-term surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A070101	Failure to Capture			
									A070909	Oversensing			
									A071206	Pacing Asynchronously			
							F26	No Health Consequences or Impact	A070101	Failure to Capture			
									A070909	Oversensing			
									A071206	Pacing Asynchronously			
					CT449	Generators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070101	Failure to Capture			
									A070909	Oversensing			
									A071206	Pacing Asynchronously			
							F26	No Health Consequences or Impact	A070101	Failure to Capture			
									A070909	Oversensing			
									A071206	Pacing Asynchronously			
					CT981	Single use	E2403	No Clinical Signs, Symptoms or Conditions	A070101	Failure to Capture			
									A070909	Oversensing			
									A071206	Pacing Asynchronously			
							F26	No Health Consequences or Impact	A070101	Failure to Capture			
									A070909	Oversensing			
									A071206	Pacing Asynchronously			
					CT983	Surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A070101	Failure to Capture			
									A070909	Oversensing			
									A071206	Pacing Asynchronously			
							F26	No Health Consequences or Impact	A070101	Failure to Capture			
									A070909	Oversensing			
									A071206	Pacing Asynchronously			
11826	2019	All other reportable incidents	47263	Cardiac resynchronization therapy implantable pacemaker	CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery			
					CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery			
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery			
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery			
11827	2021	All other reportable incidents	47265	Dual-chamber implantable pacemaker, rate-responsive	CT141	Electrodes/Leads and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	25/10/2021	25/10/2021
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	25/10/2021	25/10/2021
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	25/10/2021	25/10/2021
					CT145	Cardiovascular devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	25/10/2021	25/10/2021
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	25/10/2021	25/10/2021
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	25/10/2021	25/10/2021
					CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	25/10/2021	25/10/2021
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	25/10/2021	25/10/2021
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	25/10/2021	25/10/2021
					CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	25/10/2021	25/10/2021
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	25/10/2021	25/10/2021
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	25/10/2021	25/10/2021
					CT233	Surgical	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	25/10/2021	25/10/2021
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	25/10/2021	25/10/2021
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	25/10/2021	25/10/2021
					CT2406	Implantable	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	25/10/2021	25/10/2021
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	25/10/2021	25/10/2021
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	25/10/2021	25/10/2021
					CT2407	Active implantable	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	25/10/2021	25/10/2021
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	25/10/2021	25/10/2021
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	25/10/2021	25/10/2021
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	25/10/2021	25/10/2021
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	25/10/2021	25/10/2021
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	25/10/2021	25/10/2021
					CT2727	Active implant devices/ancillaries	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	25/10/2021	25/10/2021
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	25/10/2021	25/10/2021
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	25/10/2021	25/10/2021
					CT2750	Implantable cardiac electrical stimulators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	25/10/2021	25/10/2021
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	25/10/2021	25/10/2021
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	25/10/2021	25/10/2021
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	25/10/2021	25/10/2021
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	25/10/2021	25/10/2021
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	25/10/2021	25/10/2021
					CT2985	Active	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	25/10/2021	25/10/2021
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	25/10/2021	25/10/2021
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	25/10/2021	25/10/2021
					CT3001	Implantable metal	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	25/10/2021	25/10/2021
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	25/10/2021	25/10/2021
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	25/10/2021	25/10/2021

					CT3002	Implantable synthetic polymer	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	25/10/2021	25/10/2021
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	25/10/2021	25/10/2021
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	25/10/2021	25/10/2021
					CT3006	Implantable inorganic material	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	25/10/2021	25/10/2021
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	25/10/2021	25/10/2021
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	25/10/2021	25/10/2021
					CT314	Electrically powered	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	25/10/2021	25/10/2021
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	25/10/2021	25/10/2021
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	25/10/2021	25/10/2021
					CT321	Long-term surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	25/10/2021	25/10/2021
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	25/10/2021	25/10/2021
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	25/10/2021	25/10/2021
					CT449	Generators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	25/10/2021	25/10/2021
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	25/10/2021	25/10/2021
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	25/10/2021	25/10/2021
					CT981	Single use	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	25/10/2021	25/10/2021
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	25/10/2021	25/10/2021
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	25/10/2021	25/10/2021
					CT983	Surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	25/10/2021	25/10/2021
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	25/10/2021	25/10/2021
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	25/10/2021	25/10/2021
11828	2022	All other reportable incidents	17882	Physiologic-monitoring defibrillation system	CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A07	Electrical /Electronic Property Problem		
							F27	Problem Identified During Non-Clinical Procedure	A07	Electrical /Electronic Property Problem		
11829	2020	All other reportable incidents	47265	Dual-chamber implantable pacemaker, rate-responsive	CT141	Electrodes/Leads and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	02/08/2020	02/08/2020
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	02/08/2020	02/08/2020
							F12	Serious Injury/ Illness/ Impairment	A070504	Premature Discharge of Battery	02/08/2020	02/08/2020
							F19	Surgical Intervention	A070504	Premature Discharge of Battery	02/08/2020	02/08/2020
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	02/08/2020	02/08/2020
					CT145	Cardiovascular devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	02/08/2020	02/08/2020
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	02/08/2020	02/08/2020
							F12	Serious Injury/ Illness/ Impairment	A070504	Premature Discharge of Battery	02/08/2020	02/08/2020
							F19	Surgical Intervention	A070504	Premature Discharge of Battery	02/08/2020	02/08/2020
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	02/08/2020	02/08/2020
					CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	02/08/2020	02/08/2020
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	02/08/2020	02/08/2020
							F12	Serious Injury/ Illness/ Impairment	A070504	Premature Discharge of Battery	02/08/2020	02/08/2020
							F19	Surgical Intervention	A070504	Premature Discharge of Battery	02/08/2020	02/08/2020
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	02/08/2020	02/08/2020
					CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	02/08/2020	02/08/2020
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	02/08/2020	02/08/2020
							F12	Serious Injury/ Illness/ Impairment	A070504	Premature Discharge of Battery	02/08/2020	02/08/2020
							F19	Surgical Intervention	A070504	Premature Discharge of Battery	02/08/2020	02/08/2020
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	02/08/2020	02/08/2020
					CT233	Surgical	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	02/08/2020	02/08/2020
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	02/08/2020	02/08/2020
							F12	Serious Injury/ Illness/ Impairment	A070504	Premature Discharge of Battery	02/08/2020	02/08/2020
							F19	Surgical Intervention	A070504	Premature Discharge of Battery	02/08/2020	02/08/2020
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	02/08/2020	02/08/2020
					CT2406	Implantable	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	02/08/2020	02/08/2020
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	02/08/2020	02/08/2020
							F12	Serious Injury/ Illness/ Impairment	A070504	Premature Discharge of Battery	02/08/2020	02/08/2020
							F19	Surgical Intervention	A070504	Premature Discharge of Battery	02/08/2020	02/08/2020
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	02/08/2020	02/08/2020
					CT2407	Active implantable	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	02/08/2020	02/08/2020
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	02/08/2020	02/08/2020
							F12	Serious Injury/ Illness/ Impairment	A070504	Premature Discharge of Battery	02/08/2020	02/08/2020
							F19	Surgical Intervention	A070504	Premature Discharge of Battery	02/08/2020	02/08/2020
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	02/08/2020	02/08/2020
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	02/08/2020	02/08/2020
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	02/08/2020	02/08/2020
							F12	Serious Injury/ Illness/ Impairment	A070504	Premature Discharge of Battery	02/08/2020	02/08/2020
							F19	Surgical Intervention	A070504	Premature Discharge of Battery	02/08/2020	02/08/2020
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	02/08/2020	02/08/2020
					CT2727	Active implant devices/ancillaries	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	02/08/2020	02/08/2020
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	02/08/2020	02/08/2020
							F12	Serious Injury/ Illness/ Impairment	A070504	Premature Discharge of Battery	02/08/2020	02/08/2020
							F19	Surgical Intervention	A070504	Premature Discharge of Battery	02/08/2020	02/08/2020
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	02/08/2020	02/08/2020
					CT2750	Implantable cardiac electrical stimulators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	02/08/2020	02/08/2020
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	02/08/2020	02/08/2020
							F12	Serious Injury/ Illness/ Impairment	A070504	Premature Discharge of Battery	02/08/2020	02/08/2020
							F19	Surgical Intervention	A070504	Premature Discharge of Battery	02/08/2020	02/08/2020
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	02/08/2020	02/08/2020
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	02/08/2020	02/08/2020
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	02/08/2020	02/08/2020
							F12	Serious Injury/ Illness/ Impairment	A070504	Premature Discharge of Battery	02/08/2020	02/08/2020
							F19	Surgical Intervention	A070504	Premature Discharge of Battery	02/08/2020	02/08/2020
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	02/08/2020	02/08/2020
					CT2985	Active	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	02/08/2020	02/08/2020
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	02/08/2020	02/08/2020
							F12	Serious Injury/ Illness/ Impairment	A070504	Premature Discharge of Battery	02/08/2020	02/08/2020
							F19	Surgical Intervention	A070504	Premature Discharge of Battery	02/08/2020	02/08/2020
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	02/08/2020	02/08/2020
					CT3001	Implantable metal	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	02/08/2020	02/08/2020
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	02/08/2020	02/08/2020
							F12	Serious Injury/ Illness/ Impairment	A070504	Premature Discharge of Battery	02/08/2020	02/08/2020
							F19	Surgical Intervention	A070504	Premature Discharge of Battery	02/08/2020	02/08/2020
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	02/08/2020	02/08/2020

					CT3002	Implantable synthetic polymer	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	02/08/2020	02/08/2020
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	02/08/2020	02/08/2020
							F12	Serious Injury/ Illness/ Impairment	A070504	Premature Discharge of Battery	02/08/2020	02/08/2020
							F19	Surgical Intervention	A070504	Premature Discharge of Battery	02/08/2020	02/08/2020
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	02/08/2020	02/08/2020
					CT3006	Implantable inorganic material	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	02/08/2020	02/08/2020
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	02/08/2020	02/08/2020
							F12	Serious Injury/ Illness/ Impairment	A070504	Premature Discharge of Battery	02/08/2020	02/08/2020
							F19	Surgical Intervention	A070504	Premature Discharge of Battery	02/08/2020	02/08/2020
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	02/08/2020	02/08/2020
					CT314	Electrically powered	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	02/08/2020	02/08/2020
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	02/08/2020	02/08/2020
							F12	Serious Injury/ Illness/ Impairment	A070504	Premature Discharge of Battery	02/08/2020	02/08/2020
							F19	Surgical Intervention	A070504	Premature Discharge of Battery	02/08/2020	02/08/2020
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	02/08/2020	02/08/2020
					CT321	Long-term surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	02/08/2020	02/08/2020
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	02/08/2020	02/08/2020
							F12	Serious Injury/ Illness/ Impairment	A070504	Premature Discharge of Battery	02/08/2020	02/08/2020
							F19	Surgical Intervention	A070504	Premature Discharge of Battery	02/08/2020	02/08/2020
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	02/08/2020	02/08/2020
					CT449	Generators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	02/08/2020	02/08/2020
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	02/08/2020	02/08/2020
							F12	Serious Injury/ Illness/ Impairment	A070504	Premature Discharge of Battery	02/08/2020	02/08/2020
							F19	Surgical Intervention	A070504	Premature Discharge of Battery	02/08/2020	02/08/2020
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	02/08/2020	02/08/2020
					CT981	Single use	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	02/08/2020	02/08/2020
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	02/08/2020	02/08/2020
							F12	Serious Injury/ Illness/ Impairment	A070504	Premature Discharge of Battery	02/08/2020	02/08/2020
							F19	Surgical Intervention	A070504	Premature Discharge of Battery	02/08/2020	02/08/2020
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	02/08/2020	02/08/2020
					CT983	Surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	02/08/2020	02/08/2020
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	02/08/2020	02/08/2020
							F12	Serious Injury/ Illness/ Impairment	A070504	Premature Discharge of Battery	02/08/2020	02/08/2020
							F19	Surgical Intervention	A070504	Premature Discharge of Battery	02/08/2020	02/08/2020
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	02/08/2020	02/08/2020
11830	2019	All other reportable incidents	47265	Dual-chamber implantable pacemaker, rate-responsive	CT141	Electrodes/Leads and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
					CT145	Cardiovascular devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
					CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
					CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
					CT233	Surgical	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
					CT2406	Implantable	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
					CT2407	Active implantable	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
					CT2727	Active implant devices/ancillaries	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
					CT2750	Implantable cardiac electrical stimulators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
					CT2985	Active	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
					CT3001	Implantable metal	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
					CT3002	Implantable synthetic polymer	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
					CT3006	Implantable inorganic material	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
					CT314	Electrically powered	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
					CT321	Long-term surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
					CT449	Generators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
					CT981	Single use	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
					CT983	Surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
11831	2022	All other reportable incidents	37265	Dual-chamber implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A070910	Undersensing		
									A0712	Pacing Problem		
							F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A070910	Undersensing		
									A0712	Pacing Problem		
11832	2020	All other reportable incidents	47267	Single-chamber implantable pacemaker, rate-responsive	CT141	Electrodes/Leads and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A0712	Pacing Problem		
							F1905	Device Revision or Replacement	A0712	Pacing Problem		
					CT145	Cardiovascular devices	E2403	No Clinical Signs, Symptoms or Conditions	A0712	Pacing Problem		
							F1905	Device Revision or Replacement	A0712	Pacing Problem		
					CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A0712	Pacing Problem		
							F1905	Device Revision or Replacement	A0712	Pacing Problem		
					CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A0712	Pacing Problem		
							F1905	Device Revision or Replacement	A0712	Pacing Problem		
					CT233	Surgical	E2403	No Clinical Signs, Symptoms or Conditions	A0712	Pacing Problem		
							F1905	Device Revision or Replacement	A0712	Pacing Problem		
					CT2406	Implantable	E2403	No Clinical Signs, Symptoms or Conditions	A0712	Pacing Problem		
							F1905	Device Revision or Replacement	A0712	Pacing Problem		
					CT2407	Active implantable	E2403	No Clinical Signs, Symptoms or Conditions	A0712	Pacing Problem		
							F1905	Device Revision or Replacement	A0712	Pacing Problem		
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A0712	Pacing Problem		
							F1905	Device Revision or Replacement	A0712	Pacing Problem		
					CT2727	Active implant devices/ancillaries	E2403	No Clinical Signs, Symptoms or Conditions	A0712	Pacing Problem		
							F1905	Device Revision or Replacement	A0712	Pacing Problem		
					CT2750	Implantable cardiac electrical stimulators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A0712	Pacing Problem		
							F1905	Device Revision or Replacement	A0712	Pacing Problem		
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A0712	Pacing Problem		
							F1905	Device Revision or Replacement	A0712	Pacing Problem		
					CT2985	Active	E2403	No Clinical Signs, Symptoms or Conditions	A0712	Pacing Problem		
							F1905	Device Revision or Replacement	A0712	Pacing Problem		
					CT3001	Implantable metal	E2403	No Clinical Signs, Symptoms or Conditions	A0712	Pacing Problem		
							F1905	Device Revision or Replacement	A0712	Pacing Problem		
					CT3002	Implantable synthetic polymer	E2403	No Clinical Signs, Symptoms or Conditions	A0712	Pacing Problem		
							F1905	Device Revision or Replacement	A0712	Pacing Problem		

					CT3006	Implantable inorganic material	E2403	No Clinical Signs, Symptoms or Conditions	A0712	Pacing Problem		
							F1905	Device Revision or Replacement	A0712	Pacing Problem		
					CT314	Electrically powered	E2403	No Clinical Signs, Symptoms or Conditions	A0712	Pacing Problem		
							F1905	Device Revision or Replacement	A0712	Pacing Problem		
					CT321	Long-term surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A0712	Pacing Problem		
							F1905	Device Revision or Replacement	A0712	Pacing Problem		
					CT449	Generators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A0712	Pacing Problem		
							F1905	Device Revision or Replacement	A0712	Pacing Problem		
					CT981	Single use	E2403	No Clinical Signs, Symptoms or Conditions	A0712	Pacing Problem		
							F1905	Device Revision or Replacement	A0712	Pacing Problem		
					CT983	Surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A0712	Pacing Problem		
							F1905	Device Revision or Replacement	A0712	Pacing Problem		
11833	2022	All other reportable incidents	47267	Single-chamber implantable pacemaker, rate-responsive	CT141	Electrodes/Leads and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem	08/12/2022	08/12/2022
									A070504	Premature Discharge of Battery	08/12/2022	08/12/2022
							F08	Hospitalization or Prolonged Hospitalization	A0705	Battery Problem	08/12/2022	08/12/2022
									A070504	Premature Discharge of Battery	08/12/2022	08/12/2022
							F1903	Device Explantation	A0705	Battery Problem	08/12/2022	08/12/2022
									A070504	Premature Discharge of Battery	08/12/2022	08/12/2022
					CT145	Cardiovascular devices	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem	08/12/2022	08/12/2022
									A070504	Premature Discharge of Battery	08/12/2022	08/12/2022
							F08	Hospitalization or Prolonged Hospitalization	A0705	Battery Problem	08/12/2022	08/12/2022
									A070504	Premature Discharge of Battery	08/12/2022	08/12/2022
							F1903	Device Explantation	A0705	Battery Problem	08/12/2022	08/12/2022
									A070504	Premature Discharge of Battery	08/12/2022	08/12/2022
					CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem	08/12/2022	08/12/2022
									A070504	Premature Discharge of Battery	08/12/2022	08/12/2022
							F08	Hospitalization or Prolonged Hospitalization	A0705	Battery Problem	08/12/2022	08/12/2022
									A070504	Premature Discharge of Battery	08/12/2022	08/12/2022
							F1903	Device Explantation	A0705	Battery Problem	08/12/2022	08/12/2022
									A070504	Premature Discharge of Battery	08/12/2022	08/12/2022
					CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem	08/12/2022	08/12/2022
									A070504	Premature Discharge of Battery	08/12/2022	08/12/2022
							F08	Hospitalization or Prolonged Hospitalization	A0705	Battery Problem	08/12/2022	08/12/2022
									A070504	Premature Discharge of Battery	08/12/2022	08/12/2022
							F1903	Device Explantation	A0705	Battery Problem	08/12/2022	08/12/2022
									A070504	Premature Discharge of Battery	08/12/2022	08/12/2022
					CT233	Surgical	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem	08/12/2022	08/12/2022
									A070504	Premature Discharge of Battery	08/12/2022	08/12/2022
							F08	Hospitalization or Prolonged Hospitalization	A0705	Battery Problem	08/12/2022	08/12/2022
									A070504	Premature Discharge of Battery	08/12/2022	08/12/2022
							F1903	Device Explantation	A0705	Battery Problem	08/12/2022	08/12/2022
									A070504	Premature Discharge of Battery	08/12/2022	08/12/2022
					CT2406	Implantable	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem	08/12/2022	08/12/2022
									A070504	Premature Discharge of Battery	08/12/2022	08/12/2022
							F08	Hospitalization or Prolonged Hospitalization	A0705	Battery Problem	08/12/2022	08/12/2022
									A070504	Premature Discharge of Battery	08/12/2022	08/12/2022
							F1903	Device Explantation	A0705	Battery Problem	08/12/2022	08/12/2022
									A070504	Premature Discharge of Battery	08/12/2022	08/12/2022
					CT2407	Active implantable	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem	08/12/2022	08/12/2022
									A070504	Premature Discharge of Battery	08/12/2022	08/12/2022
							F08	Hospitalization or Prolonged Hospitalization	A0705	Battery Problem	08/12/2022	08/12/2022
									A070504	Premature Discharge of Battery	08/12/2022	08/12/2022
							F1903	Device Explantation	A0705	Battery Problem	08/12/2022	08/12/2022
									A070504	Premature Discharge of Battery	08/12/2022	08/12/2022
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem	08/12/2022	08/12/2022
									A070504	Premature Discharge of Battery	08/12/2022	08/12/2022
							F08	Hospitalization or Prolonged Hospitalization	A0705	Battery Problem	08/12/2022	08/12/2022
									A070504	Premature Discharge of Battery	08/12/2022	08/12/2022
							F1903	Device Explantation	A0705	Battery Problem	08/12/2022	08/12/2022
									A070504	Premature Discharge of Battery	08/12/2022	08/12/2022
					CT2727	Active implant devices/ancillaries	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem	08/12/2022	08/12/2022
									A070504	Premature Discharge of Battery	08/12/2022	08/12/2022
							F08	Hospitalization or Prolonged Hospitalization	A0705	Battery Problem	08/12/2022	08/12/2022
									A070504	Premature Discharge of Battery	08/12/2022	08/12/2022
							F1903	Device Explantation	A0705	Battery Problem	08/12/2022	08/12/2022
									A070504	Premature Discharge of Battery	08/12/2022	08/12/2022
					CT2750	Implantable cardiac electrical stimulators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem	08/12/2022	08/12/2022
									A070504	Premature Discharge of Battery	08/12/2022	08/12/2022
							F08	Hospitalization or Prolonged Hospitalization	A0705	Battery Problem	08/12/2022	08/12/2022
									A070504	Premature Discharge of Battery	08/12/2022	08/12/2022
							F1903	Device Explantation	A0705	Battery Problem	08/12/2022	08/12/2022
									A070504	Premature Discharge of Battery	08/12/2022	08/12/2022
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem	08/12/2022	08/12/2022
									A070504	Premature Discharge of Battery	08/12/2022	08/12/2022
							F08	Hospitalization or Prolonged Hospitalization	A0705	Battery Problem	08/12/2022	08/12/2022
									A070504	Premature Discharge of Battery	08/12/2022	08/12/2022
							F1903	Device Explantation	A0705	Battery Problem	08/12/2022	08/12/2022
									A070504	Premature Discharge of Battery	08/12/2022	08/12/2022
					CT2985	Active	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem	08/12/2022	08/12/2022
									A070504	Premature Discharge of Battery	08/12/2022	08/12/2022
							F08	Hospitalization or Prolonged Hospitalization	A0705	Battery Problem	08/12/2022	08/12/2022
									A070504	Premature Discharge of Battery	08/12/2022	08/12/2022
							F1903	Device Explantation	A0705	Battery Problem	08/12/2022	08/12/2022
									A070504	Premature Discharge of Battery	08/12/2022	08/12/2022
					CT3001	Implantable metal	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem	08/12/2022	08/12/2022
									A070504	Premature Discharge of Battery	08/12/2022	08/12/2022
							F08	Hospitalization or Prolonged Hospitalization	A0705	Battery Problem	08/12/2022	08/12/2022
									A070504	Premature Discharge of Battery	08/12/2022	08/12/2022

							F1903	Device Explantation	A0705	Battery Problem	08/12/2022	08/12/2022
									A070504	Premature Discharge of Battery	08/12/2022	08/12/2022
					CT3002	Implantable synthetic polymer	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem	08/12/2022	08/12/2022
									A070504	Premature Discharge of Battery	08/12/2022	08/12/2022
							F08	Hospitalization or Prolonged Hospitalization	A0705	Battery Problem	08/12/2022	08/12/2022
									A070504	Premature Discharge of Battery	08/12/2022	08/12/2022
							F1903	Device Explantation	A0705	Battery Problem	08/12/2022	08/12/2022
									A070504	Premature Discharge of Battery	08/12/2022	08/12/2022
					CT3006	Implantable inorganic material	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem	08/12/2022	08/12/2022
									A070504	Premature Discharge of Battery	08/12/2022	08/12/2022
							F08	Hospitalization or Prolonged Hospitalization	A0705	Battery Problem	08/12/2022	08/12/2022
									A070504	Premature Discharge of Battery	08/12/2022	08/12/2022
							F1903	Device Explantation	A0705	Battery Problem	08/12/2022	08/12/2022
									A070504	Premature Discharge of Battery	08/12/2022	08/12/2022
					CT314	Electrically powered	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem	08/12/2022	08/12/2022
									A070504	Premature Discharge of Battery	08/12/2022	08/12/2022
							F08	Hospitalization or Prolonged Hospitalization	A0705	Battery Problem	08/12/2022	08/12/2022
									A070504	Premature Discharge of Battery	08/12/2022	08/12/2022
							F1903	Device Explantation	A0705	Battery Problem	08/12/2022	08/12/2022
									A070504	Premature Discharge of Battery	08/12/2022	08/12/2022
					CT321	Long-term surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem	08/12/2022	08/12/2022
									A070504	Premature Discharge of Battery	08/12/2022	08/12/2022
							F08	Hospitalization or Prolonged Hospitalization	A0705	Battery Problem	08/12/2022	08/12/2022
									A070504	Premature Discharge of Battery	08/12/2022	08/12/2022
							F1903	Device Explantation	A0705	Battery Problem	08/12/2022	08/12/2022
									A070504	Premature Discharge of Battery	08/12/2022	08/12/2022
					CT449	Generators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem	08/12/2022	08/12/2022
									A070504	Premature Discharge of Battery	08/12/2022	08/12/2022
							F08	Hospitalization or Prolonged Hospitalization	A0705	Battery Problem	08/12/2022	08/12/2022
									A070504	Premature Discharge of Battery	08/12/2022	08/12/2022
							F1903	Device Explantation	A0705	Battery Problem	08/12/2022	08/12/2022
									A070504	Premature Discharge of Battery	08/12/2022	08/12/2022
					CT981	Single use	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem	08/12/2022	08/12/2022
									A070504	Premature Discharge of Battery	08/12/2022	08/12/2022
							F08	Hospitalization or Prolonged Hospitalization	A0705	Battery Problem	08/12/2022	08/12/2022
									A070504	Premature Discharge of Battery	08/12/2022	08/12/2022
							F1903	Device Explantation	A0705	Battery Problem	08/12/2022	08/12/2022
									A070504	Premature Discharge of Battery	08/12/2022	08/12/2022
					CT983	Surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem	08/12/2022	08/12/2022
									A070504	Premature Discharge of Battery	08/12/2022	08/12/2022
							F08	Hospitalization or Prolonged Hospitalization	A0705	Battery Problem	08/12/2022	08/12/2022
									A070504	Premature Discharge of Battery	08/12/2022	08/12/2022
							F1903	Device Explantation	A0705	Battery Problem	08/12/2022	08/12/2022
									A070504	Premature Discharge of Battery	08/12/2022	08/12/2022
11834	2022	All other reportable incidents	35223	Endocardial/interventricular septal pacing lead	CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A0401	Break		
							F22	Unexpected Diagnostic Intervention	A0401	Break		
11835	2022	All other reportable incidents	37265	Dual-chamber implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
11836	2022	All other reportable incidents	47265	Dual-chamber implantable pacemaker, rate-responsive	CT141	Electrodes/Leads and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem		
							F1901	Additional Surgery	A24	No Apparent Device Problem		
					CT145	Cardiovascular devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem		
							F1901	Additional Surgery	A24	No Apparent Device Problem		
					CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem		
							F1901	Additional Surgery	A24	No Apparent Device Problem		
					CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem		
							F1901	Additional Surgery	A24	No Apparent Device Problem		
					CT233	Surgical	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem		
							F1901	Additional Surgery	A24	No Apparent Device Problem		
					CT2406	Implantable	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem		
							F1901	Additional Surgery	A24	No Apparent Device Problem		
					CT2407	Active implantable	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem		
							F1901	Additional Surgery	A24	No Apparent Device Problem		
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem		
							F1901	Additional Surgery	A24	No Apparent Device Problem		
					CT2727	Active implant devices/ancillaries	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem		
							F1901	Additional Surgery	A24	No Apparent Device Problem		
					CT2750	Implantable cardiac electrical stimulators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem		
							F1901	Additional Surgery	A24	No Apparent Device Problem		
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem		
							F1901	Additional Surgery	A24	No Apparent Device Problem		
					CT2985	Active	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem		
							F1901	Additional Surgery	A24	No Apparent Device Problem		
					CT3001	Implantable metal	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem		
							F1901	Additional Surgery	A24	No Apparent Device Problem		
					CT3002	Implantable synthetic polymer	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem		
							F1901	Additional Surgery	A24	No Apparent Device Problem		
					CT3006	Implantable inorganic material	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem		
							F1901	Additional Surgery	A24	No Apparent Device Problem		
					CT314	Electrically powered	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem		
							F1901	Additional Surgery	A24	No Apparent Device Problem		
					CT321	Long-term surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem		
							F1901	Additional Surgery	A24	No Apparent Device Problem		
					CT449	Generators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem		
							F1901	Additional Surgery	A24	No Apparent Device Problem		
					CT981	Single use	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem		
							F1901	Additional Surgery	A24	No Apparent Device Problem		
					CT983	Surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem		
							F1901	Additional Surgery	A24	No Apparent Device Problem		

11837	2022	All other reportable incidents	47265	Dual-chamber implantable pacemaker, rate-responsive	CT141	Electrodes/Leads and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem		
							F1901	Additional Surgery	A24	No Apparent Device Problem		
					CT145	Cardiovascular devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem		
							F1901	Additional Surgery	A24	No Apparent Device Problem		
					CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem		
							F1901	Additional Surgery	A24	No Apparent Device Problem		
					CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem		
							F1901	Additional Surgery	A24	No Apparent Device Problem		
					CT233	Surgical	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem		
							F1901	Additional Surgery	A24	No Apparent Device Problem		
					CT2406	Implantable	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem		
							F1901	Additional Surgery	A24	No Apparent Device Problem		
					CT2407	Active implantable	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem		
							F1901	Additional Surgery	A24	No Apparent Device Problem		
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem		
							F1901	Additional Surgery	A24	No Apparent Device Problem		
					CT2727	Active implant devices/ancillaries	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem		
							F1901	Additional Surgery	A24	No Apparent Device Problem		
					CT2750	Implantable cardiac electrical stimulators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem		
							F1901	Additional Surgery	A24	No Apparent Device Problem		
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem		
							F1901	Additional Surgery	A24	No Apparent Device Problem		
					CT2985	Active	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem		
							F1901	Additional Surgery	A24	No Apparent Device Problem		
					CT3001	Implantable metal	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem		
							F1901	Additional Surgery	A24	No Apparent Device Problem		
					CT3002	Implantable synthetic polymer	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem		
							F1901	Additional Surgery	A24	No Apparent Device Problem		
					CT3006	Implantable inorganic material	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem		
							F1901	Additional Surgery	A24	No Apparent Device Problem		
					CT314	Electrically powered	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem		
							F1901	Additional Surgery	A24	No Apparent Device Problem		
					CT321	Long-term surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem		
							F1901	Additional Surgery	A24	No Apparent Device Problem		
					CT449	Generators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem		
							F1901	Additional Surgery	A24	No Apparent Device Problem		
					CT981	Single use	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem		
							F1901	Additional Surgery	A24	No Apparent Device Problem		
					CT983	Surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem		
							F1901	Additional Surgery	A24	No Apparent Device Problem		
11838	2022	All other reportable incidents	47265	Dual-chamber implantable pacemaker, rate-responsive	CT141	Electrodes/Leads and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem		
							F1901	Additional Surgery	A24	No Apparent Device Problem		
					CT145	Cardiovascular devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem		
							F1901	Additional Surgery	A24	No Apparent Device Problem		
					CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem		
							F1901	Additional Surgery	A24	No Apparent Device Problem		
					CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem		
							F1901	Additional Surgery	A24	No Apparent Device Problem		
					CT233	Surgical	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem		
							F1901	Additional Surgery	A24	No Apparent Device Problem		
					CT2406	Implantable	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem		
							F1901	Additional Surgery	A24	No Apparent Device Problem		
					CT2407	Active implantable	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem		
							F1901	Additional Surgery	A24	No Apparent Device Problem		
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem		
							F1901	Additional Surgery	A24	No Apparent Device Problem		
					CT2727	Active implant devices/ancillaries	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem		
							F1901	Additional Surgery	A24	No Apparent Device Problem		
					CT2750	Implantable cardiac electrical stimulators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem		
							F1901	Additional Surgery	A24	No Apparent Device Problem		
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem		
							F1901	Additional Surgery	A24	No Apparent Device Problem		
					CT2985	Active	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem		
							F1901	Additional Surgery	A24	No Apparent Device Problem		
					CT3001	Implantable metal	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem		
							F1901	Additional Surgery	A24	No Apparent Device Problem		
					CT3002	Implantable synthetic polymer	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem		
							F1901	Additional Surgery	A24	No Apparent Device Problem		
					CT3006	Implantable inorganic material	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem		
							F1901	Additional Surgery	A24	No Apparent Device Problem		
					CT314	Electrically powered	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem		
							F1901	Additional Surgery	A24	No Apparent Device Problem		
					CT321	Long-term surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem		
							F1901	Additional Surgery	A24	No Apparent Device Problem		
					CT449	Generators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem		
							F1901	Additional Surgery	A24	No Apparent Device Problem		
					CT981	Single use	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem		
							F1901	Additional Surgery	A24	No Apparent Device Problem		
					CT983	Surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem		
							F1901	Additional Surgery	A24	No Apparent Device Problem		
11839	2022	All other reportable incidents	47265	Dual-chamber implantable pacemaker, rate-responsive	CT141	Electrodes/Leads and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	05/12/2022	05/12/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	05/12/2022	05/12/2022
					CT145	Cardiovascular devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	05/12/2022	05/12/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	05/12/2022	05/12/2022
					CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	05/12/2022	05/12/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	05/12/2022	05/12/2022
					CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	05/12/2022	05/12/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	05/12/2022	05/12/2022

					CT233	Surgical	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	05/12/2022	05/12/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	05/12/2022	05/12/2022
					CT2406	Implantable	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	05/12/2022	05/12/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	05/12/2022	05/12/2022
					CT2407	Active implantable	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	05/12/2022	05/12/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	05/12/2022	05/12/2022
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	05/12/2022	05/12/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	05/12/2022	05/12/2022
					CT2727	Active implant devices/ancillaries	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	05/12/2022	05/12/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	05/12/2022	05/12/2022
					CT2750	Implantable cardiac electrical stimulators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	05/12/2022	05/12/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	05/12/2022	05/12/2022
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	05/12/2022	05/12/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	05/12/2022	05/12/2022
					CT2985	Active	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	05/12/2022	05/12/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	05/12/2022	05/12/2022
					CT3001	Implantable metal	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	05/12/2022	05/12/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	05/12/2022	05/12/2022
					CT3002	Implantable synthetic polymer	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	05/12/2022	05/12/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	05/12/2022	05/12/2022
					CT3006	Implantable inorganic material	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	05/12/2022	05/12/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	05/12/2022	05/12/2022
					CT314	Electrically powered	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	05/12/2022	05/12/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	05/12/2022	05/12/2022
					CT321	Long-term surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	05/12/2022	05/12/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	05/12/2022	05/12/2022
					CT449	Generators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	05/12/2022	05/12/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	05/12/2022	05/12/2022
					CT981	Single use	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	05/12/2022	05/12/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	05/12/2022	05/12/2022
					CT983	Surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	05/12/2022	05/12/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	05/12/2022	05/12/2022
11840	2022	All other reportable incidents	63188	Left-ventricular wireless pacing system sensor/transmitter	CT145	Cardiovascular devices			A0708	Power Problem		
					CT1577	Wireless monitoring system transmitters			A0708	Power Problem		
					CT1684	Cardiac pacemakers and associated devices			A0708	Power Problem		
					CT2727	Active implant devices/ancillaries			A0708	Power Problem		
					CT2750	Implantable cardiac electrical stimulators and associated devices			A0708	Power Problem		
					CT2885	Physiological monitoring/measuring system transmitters/receivers/transceivers			A0708	Power Problem		
11841	2022	All other reportable incidents	47265	Dual-chamber implantable pacemaker, rate-responsive	CT141	Electrodes/Leads and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	07/12/2022	07/12/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	07/12/2022	07/12/2022
					CT145	Cardiovascular devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	07/12/2022	07/12/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	07/12/2022	07/12/2022
					CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	07/12/2022	07/12/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	07/12/2022	07/12/2022
					CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	07/12/2022	07/12/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	07/12/2022	07/12/2022
					CT233	Surgical	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	07/12/2022	07/12/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	07/12/2022	07/12/2022
					CT2406	Implantable	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	07/12/2022	07/12/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	07/12/2022	07/12/2022
					CT2407	Active implantable	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	07/12/2022	07/12/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	07/12/2022	07/12/2022
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	07/12/2022	07/12/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	07/12/2022	07/12/2022
					CT2727	Active implant devices/ancillaries	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	07/12/2022	07/12/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	07/12/2022	07/12/2022
					CT2750	Implantable cardiac electrical stimulators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	07/12/2022	07/12/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	07/12/2022	07/12/2022
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	07/12/2022	07/12/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	07/12/2022	07/12/2022
					CT2985	Active	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	07/12/2022	07/12/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	07/12/2022	07/12/2022
					CT3001	Implantable metal	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	07/12/2022	07/12/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	07/12/2022	07/12/2022
					CT3002	Implantable synthetic polymer	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	07/12/2022	07/12/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	07/12/2022	07/12/2022
					CT3006	Implantable inorganic material	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	07/12/2022	07/12/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	07/12/2022	07/12/2022
					CT314	Electrically powered	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	07/12/2022	07/12/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	07/12/2022	07/12/2022
					CT321	Long-term surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	07/12/2022	07/12/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	07/12/2022	07/12/2022
					CT449	Generators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	07/12/2022	07/12/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	07/12/2022	07/12/2022
					CT981	Single use	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	07/12/2022	07/12/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	07/12/2022	07/12/2022
					CT983	Surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	07/12/2022	07/12/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	07/12/2022	07/12/2022
11842	2022	All other reportable incidents	47265	Dual-chamber implantable pacemaker, rate-responsive	CT141	Electrodes/Leads and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	28/09/2022	28/09/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	28/09/2022	28/09/2022
					CT145	Cardiovascular devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	28/09/2022	28/09/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	28/09/2022	28/09/2022
					CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	28/09/2022	28/09/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	28/09/2022	28/09/2022
					CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	28/09/2022	28/09/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	28/09/2022	28/09/2022

					CT233	Surgical	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	28/09/2022	28/09/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	28/09/2022	28/09/2022
					CT2406	Implantable	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	28/09/2022	28/09/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	28/09/2022	28/09/2022
					CT2407	Active implantable	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	28/09/2022	28/09/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	28/09/2022	28/09/2022
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	28/09/2022	28/09/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	28/09/2022	28/09/2022
					CT2727	Active implant devices/ancillaries	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	28/09/2022	28/09/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	28/09/2022	28/09/2022
					CT2750	Implantable cardiac electrical stimulators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	28/09/2022	28/09/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	28/09/2022	28/09/2022
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	28/09/2022	28/09/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	28/09/2022	28/09/2022
					CT2985	Active	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	28/09/2022	28/09/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	28/09/2022	28/09/2022
					CT3001	Implantable metal	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	28/09/2022	28/09/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	28/09/2022	28/09/2022
					CT3002	Implantable synthetic polymer	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	28/09/2022	28/09/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	28/09/2022	28/09/2022
					CT3006	Implantable inorganic material	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	28/09/2022	28/09/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	28/09/2022	28/09/2022
					CT314	Electrically powered	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	28/09/2022	28/09/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	28/09/2022	28/09/2022
					CT321	Long-term surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	28/09/2022	28/09/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	28/09/2022	28/09/2022
					CT449	Generators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	28/09/2022	28/09/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	28/09/2022	28/09/2022
					CT981	Single use	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	28/09/2022	28/09/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	28/09/2022	28/09/2022
					CT983	Surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	28/09/2022	28/09/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	28/09/2022	28/09/2022
11843	2022		47265	Dual-chamber implantable pacemaker, rate-responsive	CT141	Electrodes/Leads and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	10/12/2022	10/12/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	10/12/2022	10/12/2022
					CT145	Cardiovascular devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	10/12/2022	10/12/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	10/12/2022	10/12/2022
					CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	10/12/2022	10/12/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	10/12/2022	10/12/2022
					CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	10/12/2022	10/12/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	10/12/2022	10/12/2022
					CT233	Surgical	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	10/12/2022	10/12/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	10/12/2022	10/12/2022
					CT2406	Implantable	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	10/12/2022	10/12/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	10/12/2022	10/12/2022
					CT2407	Active implantable	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	10/12/2022	10/12/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	10/12/2022	10/12/2022
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	10/12/2022	10/12/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	10/12/2022	10/12/2022
					CT2727	Active implant devices/ancillaries	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	10/12/2022	10/12/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	10/12/2022	10/12/2022
					CT2750	Implantable cardiac electrical stimulators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	10/12/2022	10/12/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	10/12/2022	10/12/2022
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	10/12/2022	10/12/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	10/12/2022	10/12/2022
					CT2985	Active	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	10/12/2022	10/12/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	10/12/2022	10/12/2022
					CT3001	Implantable metal	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	10/12/2022	10/12/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	10/12/2022	10/12/2022
					CT3002	Implantable synthetic polymer	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	10/12/2022	10/12/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	10/12/2022	10/12/2022
					CT3006	Implantable inorganic material	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	10/12/2022	10/12/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	10/12/2022	10/12/2022
					CT314	Electrically powered	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	10/12/2022	10/12/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	10/12/2022	10/12/2022
					CT321	Long-term surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	10/12/2022	10/12/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	10/12/2022	10/12/2022
					CT449	Generators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	10/12/2022	10/12/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	10/12/2022	10/12/2022
					CT981	Single use	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	10/12/2022	10/12/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	10/12/2022	10/12/2022
					CT983	Surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	10/12/2022	10/12/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	10/12/2022	10/12/2022
11844	2022	All other reportable incidents	47265	Dual-chamber implantable pacemaker, rate-responsive	CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	11/12/2022	11/12/2022
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	11/12/2022	11/12/2022
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	11/12/2022	11/12/2022
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	11/12/2022	11/12/2022
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	11/12/2022	11/12/2022
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	11/12/2022	11/12/2022
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	11/12/2022	11/12/2022
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	11/12/2022	11/12/2022
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	11/12/2022	11/12/2022
11845	2022	All other reportable incidents	47265	Dual-chamber implantable pacemaker, rate-responsive	CT141	Electrodes/Leads and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	11/12/2022	11/12/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	11/12/2022	11/12/2022
					CT145	Cardiovascular devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	11/12/2022	11/12/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	11/12/2022	11/12/2022
					CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	11/12/2022	11/12/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	11/12/2022	11/12/2022
					CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	11/12/2022	11/12/2022

							F1901	Additional Surgery	A24	No Apparent Device Problem	11/12/2022	11/12/2022
					CT233	Surgical	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	11/12/2022	11/12/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	11/12/2022	11/12/2022
					CT2406	Implantable	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	11/12/2022	11/12/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	11/12/2022	11/12/2022
					CT2407	Active implantable	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	11/12/2022	11/12/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	11/12/2022	11/12/2022
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	11/12/2022	11/12/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	11/12/2022	11/12/2022
					CT2727	Active implant devices/ancillaries	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	11/12/2022	11/12/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	11/12/2022	11/12/2022
					CT2750	Implantable cardiac electrical stimulators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	11/12/2022	11/12/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	11/12/2022	11/12/2022
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	11/12/2022	11/12/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	11/12/2022	11/12/2022
					CT2985	Active	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	11/12/2022	11/12/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	11/12/2022	11/12/2022
					CT3001	Implantable metal	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	11/12/2022	11/12/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	11/12/2022	11/12/2022
					CT3002	Implantable synthetic polymer	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	11/12/2022	11/12/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	11/12/2022	11/12/2022
					CT3006	Implantable inorganic material	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	11/12/2022	11/12/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	11/12/2022	11/12/2022
					CT314	Electrically powered	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	11/12/2022	11/12/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	11/12/2022	11/12/2022
					CT321	Long-term surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	11/12/2022	11/12/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	11/12/2022	11/12/2022
					CT449	Generators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	11/12/2022	11/12/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	11/12/2022	11/12/2022
					CT981	Single use	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	11/12/2022	11/12/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	11/12/2022	11/12/2022
					CT983	Surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	11/12/2022	11/12/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	11/12/2022	11/12/2022
11846	2022	All other reportable incidents	17882	Physiologic-monitoring defibrillation system	CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A0801	Failure to Calibrate		
							F27	Problem Identified During Non-Clinical Procedure	A0801	Failure to Calibrate		
11847	2022		47265	Dual-chamber implantable pacemaker, rate-responsive	CT141	Electrodes/Leads and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	01/10/2022	31/10/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	01/10/2022	31/10/2022
					CT145	Cardiovascular devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	01/10/2022	31/10/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	01/10/2022	31/10/2022
					CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	01/10/2022	31/10/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	01/10/2022	31/10/2022
					CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	01/10/2022	31/10/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	01/10/2022	31/10/2022
					CT233	Surgical	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	01/10/2022	31/10/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	01/10/2022	31/10/2022
					CT2406	Implantable	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	01/10/2022	31/10/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	01/10/2022	31/10/2022
					CT2407	Active implantable	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	01/10/2022	31/10/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	01/10/2022	31/10/2022
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	01/10/2022	31/10/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	01/10/2022	31/10/2022
					CT2727	Active implant devices/ancillaries	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	01/10/2022	31/10/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	01/10/2022	31/10/2022
					CT2750	Implantable cardiac electrical stimulators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	01/10/2022	31/10/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	01/10/2022	31/10/2022
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	01/10/2022	31/10/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	01/10/2022	31/10/2022
					CT2985	Active	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	01/10/2022	31/10/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	01/10/2022	31/10/2022
					CT3001	Implantable metal	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	01/10/2022	31/10/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	01/10/2022	31/10/2022
					CT3002	Implantable synthetic polymer	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	01/10/2022	31/10/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	01/10/2022	31/10/2022
					CT3006	Implantable inorganic material	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	01/10/2022	31/10/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	01/10/2022	31/10/2022
					CT314	Electrically powered	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	01/10/2022	31/10/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	01/10/2022	31/10/2022
					CT321	Long-term surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	01/10/2022	31/10/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	01/10/2022	31/10/2022
					CT449	Generators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	01/10/2022	31/10/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	01/10/2022	31/10/2022
					CT981	Single use	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	01/10/2022	31/10/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	01/10/2022	31/10/2022
					CT983	Surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	01/10/2022	31/10/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	01/10/2022	31/10/2022
11848	2022		47265	Dual-chamber implantable pacemaker, rate-responsive	CT141	Electrodes/Leads and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	27/09/2022	27/09/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	27/09/2022	27/09/2022
					CT145	Cardiovascular devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	27/09/2022	27/09/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	27/09/2022	27/09/2022
					CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	27/09/2022	27/09/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	27/09/2022	27/09/2022
					CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	27/09/2022	27/09/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	27/09/2022	27/09/2022
					CT233	Surgical	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	27/09/2022	27/09/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	27/09/2022	27/09/2022
					CT2406	Implantable	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	27/09/2022	27/09/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	27/09/2022	27/09/2022
					CT2407	Active implantable	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	27/09/2022	27/09/2022

							F1901	Additional Surgery	A24	No Apparent Device Problem	27/09/2022	27/09/2022
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	27/09/2022	27/09/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	27/09/2022	27/09/2022
					CT2727	Active implant devices/ancillaries	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	27/09/2022	27/09/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	27/09/2022	27/09/2022
					CT2750	Implantable cardiac electrical stimulators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	27/09/2022	27/09/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	27/09/2022	27/09/2022
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	27/09/2022	27/09/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	27/09/2022	27/09/2022
					CT2985	Active	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	27/09/2022	27/09/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	27/09/2022	27/09/2022
					CT3001	Implantable metal	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	27/09/2022	27/09/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	27/09/2022	27/09/2022
					CT3002	Implantable synthetic polymer	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	27/09/2022	27/09/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	27/09/2022	27/09/2022
					CT3006	Implantable inorganic material	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	27/09/2022	27/09/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	27/09/2022	27/09/2022
					CT314	Electrically powered	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	27/09/2022	27/09/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	27/09/2022	27/09/2022
					CT321	Long-term surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	27/09/2022	27/09/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	27/09/2022	27/09/2022
					CT449	Generators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	27/09/2022	27/09/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	27/09/2022	27/09/2022
					CT981	Single use	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	27/09/2022	27/09/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	27/09/2022	27/09/2022
					CT983	Surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	27/09/2022	27/09/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	27/09/2022	27/09/2022
11849	2022	All other reportable incidents	47263	Cardiac resynchronization therapy implantable pacemaker	CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A0908	Incorrect, Inadequate or Imprecise Result or Readings	08/12/2022	08/12/2022
							F08	Hospitalization or Prolonged Hospitalization	A0908	Incorrect, Inadequate or Imprecise Result or Readings	08/12/2022	08/12/2022
							F1905	Device Revision or Replacement	A0908	Incorrect, Inadequate or Imprecise Result or Readings	08/12/2022	08/12/2022
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A0908	Incorrect, Inadequate or Imprecise Result or Readings	08/12/2022	08/12/2022
							F08	Hospitalization or Prolonged Hospitalization	A0908	Incorrect, Inadequate or Imprecise Result or Readings	08/12/2022	08/12/2022
							F1905	Device Revision or Replacement	A0908	Incorrect, Inadequate or Imprecise Result or Readings	08/12/2022	08/12/2022
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A0908	Incorrect, Inadequate or Imprecise Result or Readings	08/12/2022	08/12/2022
							F08	Hospitalization or Prolonged Hospitalization	A0908	Incorrect, Inadequate or Imprecise Result or Readings	08/12/2022	08/12/2022
							F1905	Device Revision or Replacement	A0908	Incorrect, Inadequate or Imprecise Result or Readings	08/12/2022	08/12/2022
11850	2022	All other reportable incidents	47265	Dual-chamber implantable pacemaker, rate-responsive	CT141	Electrodes/Leads and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem		
							F1901	Additional Surgery	A24	No Apparent Device Problem		
					CT145	Cardiovascular devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem		
							F1901	Additional Surgery	A24	No Apparent Device Problem		
					CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem		
							F1901	Additional Surgery	A24	No Apparent Device Problem		
					CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem		
							F1901	Additional Surgery	A24	No Apparent Device Problem		
					CT233	Surgical	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem		
							F1901	Additional Surgery	A24	No Apparent Device Problem		
					CT2406	Implantable	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem		
							F1901	Additional Surgery	A24	No Apparent Device Problem		
					CT2407	Active implantable	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem		
							F1901	Additional Surgery	A24	No Apparent Device Problem		
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem		
							F1901	Additional Surgery	A24	No Apparent Device Problem		
					CT2727	Active implant devices/ancillaries	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem		
							F1901	Additional Surgery	A24	No Apparent Device Problem		
					CT2750	Implantable cardiac electrical stimulators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem		
							F1901	Additional Surgery	A24	No Apparent Device Problem		
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem		
							F1901	Additional Surgery	A24	No Apparent Device Problem		
					CT2985	Active	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem		
							F1901	Additional Surgery	A24	No Apparent Device Problem		
					CT3001	Implantable metal	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem		
							F1901	Additional Surgery	A24	No Apparent Device Problem		
					CT3002	Implantable synthetic polymer	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem		
							F1901	Additional Surgery	A24	No Apparent Device Problem		
					CT3006	Implantable inorganic material	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem		
							F1901	Additional Surgery	A24	No Apparent Device Problem		
					CT314	Electrically powered	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem		
							F1901	Additional Surgery	A24	No Apparent Device Problem		
					CT321	Long-term surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem		
							F1901	Additional Surgery	A24	No Apparent Device Problem		
					CT449	Generators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem		
							F1901	Additional Surgery	A24	No Apparent Device Problem		
					CT981	Single use	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem		
							F1901	Additional Surgery	A24	No Apparent Device Problem		
					CT983	Surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem		
							F1901	Additional Surgery	A24	No Apparent Device Problem		
11851	2022	All other reportable incidents	47270	Cardiac resynchronization therapy implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E2311	Discomfort	A070909	Oversensing		
									A070912	Incorrect Interpretation of Signal		
									A071302	Inappropriate/Inadequate Shock/Stimulation		
							F12	Serious Injury/ Illness/ Impairment	A070909	Oversensing		
									A070912	Incorrect Interpretation of Signal		

									A071302	Inappropriate/Inadequate Shock/Stimulation		
11852	2022		47265	Dual-chamber implantable pacemaker, rate-responsive	CT141	Electrodes/Leads and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem		
							F1901	Additional Surgery	A24	No Apparent Device Problem		
					CT145	Cardiovascular devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem		
							F1901	Additional Surgery	A24	No Apparent Device Problem		
					CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem		
							F1901	Additional Surgery	A24	No Apparent Device Problem		
					CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem		
							F1901	Additional Surgery	A24	No Apparent Device Problem		
					CT233	Surgical	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem		
							F1901	Additional Surgery	A24	No Apparent Device Problem		
					CT2406	Implantable	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem		
							F1901	Additional Surgery	A24	No Apparent Device Problem		
					CT2407	Active implantable	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem		
							F1901	Additional Surgery	A24	No Apparent Device Problem		
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem		
							F1901	Additional Surgery	A24	No Apparent Device Problem		
					CT2727	Active implant devices/ancillaries	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem		
							F1901	Additional Surgery	A24	No Apparent Device Problem		
					CT2750	Implantable cardiac electrical stimulators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem		
							F1901	Additional Surgery	A24	No Apparent Device Problem		
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem		
							F1901	Additional Surgery	A24	No Apparent Device Problem		
					CT2985	Active	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem		
							F1901	Additional Surgery	A24	No Apparent Device Problem		
					CT3001	Implantable metal	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem		
							F1901	Additional Surgery	A24	No Apparent Device Problem		
					CT3002	Implantable synthetic polymer	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem		
							F1901	Additional Surgery	A24	No Apparent Device Problem		
					CT3006	Implantable inorganic material	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem		
							F1901	Additional Surgery	A24	No Apparent Device Problem		
					CT314	Electrically powered	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem		
							F1901	Additional Surgery	A24	No Apparent Device Problem		
					CT321	Long-term surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem		
							F1901	Additional Surgery	A24	No Apparent Device Problem		
					CT449	Generators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem		
							F1901	Additional Surgery	A24	No Apparent Device Problem		
					CT981	Single use	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem		
							F1901	Additional Surgery	A24	No Apparent Device Problem		
					CT983	Surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem		
							F1901	Additional Surgery	A24	No Apparent Device Problem		
11853	2022		47265	Dual-chamber implantable pacemaker, rate-responsive	CT141	Electrodes/Leads and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	05/12/2022	05/12/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	05/12/2022	05/12/2022
					CT145	Cardiovascular devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	05/12/2022	05/12/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	05/12/2022	05/12/2022
					CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	05/12/2022	05/12/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	05/12/2022	05/12/2022
					CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	05/12/2022	05/12/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	05/12/2022	05/12/2022
					CT233	Surgical	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	05/12/2022	05/12/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	05/12/2022	05/12/2022
					CT2406	Implantable	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	05/12/2022	05/12/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	05/12/2022	05/12/2022
					CT2407	Active implantable	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	05/12/2022	05/12/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	05/12/2022	05/12/2022
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	05/12/2022	05/12/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	05/12/2022	05/12/2022
					CT2727	Active implant devices/ancillaries	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	05/12/2022	05/12/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	05/12/2022	05/12/2022
					CT2750	Implantable cardiac electrical stimulators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	05/12/2022	05/12/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	05/12/2022	05/12/2022
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	05/12/2022	05/12/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	05/12/2022	05/12/2022
					CT2985	Active	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	05/12/2022	05/12/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	05/12/2022	05/12/2022
					CT3001	Implantable metal	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	05/12/2022	05/12/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	05/12/2022	05/12/2022
					CT3002	Implantable synthetic polymer	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	05/12/2022	05/12/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	05/12/2022	05/12/2022
					CT3006	Implantable inorganic material	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	05/12/2022	05/12/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	05/12/2022	05/12/2022
					CT314	Electrically powered	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	05/12/2022	05/12/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	05/12/2022	05/12/2022
					CT321	Long-term surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	05/12/2022	05/12/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	05/12/2022	05/12/2022
					CT449	Generators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	05/12/2022	05/12/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	05/12/2022	05/12/2022
					CT981	Single use	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	05/12/2022	05/12/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	05/12/2022	05/12/2022
					CT983	Surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	05/12/2022	05/12/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	05/12/2022	05/12/2022
11854	2022	All other reportable incidents	47265	Dual-chamber implantable pacemaker, rate-responsive	CT141	Electrodes/Leads and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	05/12/2022	05/12/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	05/12/2022	05/12/2022
					CT145	Cardiovascular devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	05/12/2022	05/12/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	05/12/2022	05/12/2022
					CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	05/12/2022	05/12/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	05/12/2022	05/12/2022
					CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	05/12/2022	05/12/2022

							F1901	Additional Surgery	A24	No Apparent Device Problem	05/12/2022	05/12/2022
					CT233	Surgical	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	05/12/2022	05/12/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	05/12/2022	05/12/2022
					CT2406	Implantable	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	05/12/2022	05/12/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	05/12/2022	05/12/2022
					CT2407	Active implantable	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	05/12/2022	05/12/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	05/12/2022	05/12/2022
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	05/12/2022	05/12/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	05/12/2022	05/12/2022
					CT2727	Active implant devices/ancillaries	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	05/12/2022	05/12/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	05/12/2022	05/12/2022
					CT2750	Implantable cardiac electrical stimulators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	05/12/2022	05/12/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	05/12/2022	05/12/2022
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	05/12/2022	05/12/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	05/12/2022	05/12/2022
					CT2985	Active	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	05/12/2022	05/12/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	05/12/2022	05/12/2022
					CT3001	Implantable metal	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	05/12/2022	05/12/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	05/12/2022	05/12/2022
					CT3002	Implantable synthetic polymer	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	05/12/2022	05/12/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	05/12/2022	05/12/2022
					CT3006	Implantable inorganic material	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	05/12/2022	05/12/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	05/12/2022	05/12/2022
					CT314	Electrically powered	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	05/12/2022	05/12/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	05/12/2022	05/12/2022
					CT321	Long-term surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	05/12/2022	05/12/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	05/12/2022	05/12/2022
					CT449	Generators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	05/12/2022	05/12/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	05/12/2022	05/12/2022
					CT981	Single use	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	05/12/2022	05/12/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	05/12/2022	05/12/2022
					CT983	Surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	05/12/2022	05/12/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	05/12/2022	05/12/2022
11855	2022	All other reportable incidents	47265	Dual-chamber implantable pacemaker, rate-responsive	CT1685	Cardiac pacemakers	E011902	Presyncope	A0712	Pacing Problem		
							F12	Serious Injury/ Illness/ Impairment	A0712	Pacing Problem		
							F1905	Device Revision or Replacement	A0712	Pacing Problem		
							F23	Unexpected Medical Intervention	A0712	Pacing Problem		
					CT257	Implantable cardiac pacemakers	E011902	Presyncope	A0712	Pacing Problem		
							F12	Serious Injury/ Illness/ Impairment	A0712	Pacing Problem		
							F1905	Device Revision or Replacement	A0712	Pacing Problem		
							F23	Unexpected Medical Intervention	A0712	Pacing Problem		
					CT2866	Implantable cardiac electrical stimulators	E011902	Presyncope	A0712	Pacing Problem		
							F12	Serious Injury/ Illness/ Impairment	A0712	Pacing Problem		
							F1905	Device Revision or Replacement	A0712	Pacing Problem		
							F23	Unexpected Medical Intervention	A0712	Pacing Problem		
11856	2022	All other reportable incidents	47265	Dual-chamber implantable pacemaker, rate-responsive	CT141	Electrodes/Leads and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem		
							F1901	Additional Surgery	A24	No Apparent Device Problem		
					CT145	Cardiovascular devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem		
							F1901	Additional Surgery	A24	No Apparent Device Problem		
					CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem		
							F1901	Additional Surgery	A24	No Apparent Device Problem		
					CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem		
							F1901	Additional Surgery	A24	No Apparent Device Problem		
					CT233	Surgical	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem		
							F1901	Additional Surgery	A24	No Apparent Device Problem		
					CT2406	Implantable	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem		
							F1901	Additional Surgery	A24	No Apparent Device Problem		
					CT2407	Active implantable	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem		
							F1901	Additional Surgery	A24	No Apparent Device Problem		
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem		
							F1901	Additional Surgery	A24	No Apparent Device Problem		
					CT2727	Active implant devices/ancillaries	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem		
							F1901	Additional Surgery	A24	No Apparent Device Problem		
					CT2750	Implantable cardiac electrical stimulators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem		
							F1901	Additional Surgery	A24	No Apparent Device Problem		
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem		
							F1901	Additional Surgery	A24	No Apparent Device Problem		
					CT2985	Active	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem		
							F1901	Additional Surgery	A24	No Apparent Device Problem		
					CT3001	Implantable metal	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem		
							F1901	Additional Surgery	A24	No Apparent Device Problem		
					CT3002	Implantable synthetic polymer	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem		
							F1901	Additional Surgery	A24	No Apparent Device Problem		
					CT3006	Implantable inorganic material	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem		
							F1901	Additional Surgery	A24	No Apparent Device Problem		
					CT314	Electrically powered	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem		
							F1901	Additional Surgery	A24	No Apparent Device Problem		
					CT321	Long-term surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem		
							F1901	Additional Surgery	A24	No Apparent Device Problem		
					CT449	Generators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem		
							F1901	Additional Surgery	A24	No Apparent Device Problem		
					CT981	Single use	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem		
							F1901	Additional Surgery	A24	No Apparent Device Problem		
					CT983	Surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem		
							F1901	Additional Surgery	A24	No Apparent Device Problem		
11857	2022	All other reportable incidents	47265	Dual-chamber implantable pacemaker, rate-responsive	CT141	Electrodes/Leads and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	29/11/2022	29/11/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	29/11/2022	29/11/2022
					CT145	Cardiovascular devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	29/11/2022	29/11/2022

							F1901	Additional Surgery	A24	No Apparent Device Problem	29/11/2022	29/11/2022
					CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	29/11/2022	29/11/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	29/11/2022	29/11/2022
					CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	29/11/2022	29/11/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	29/11/2022	29/11/2022
					CT233	Surgical	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	29/11/2022	29/11/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	29/11/2022	29/11/2022
					CT2406	Implantable	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	29/11/2022	29/11/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	29/11/2022	29/11/2022
					CT2407	Active implantable	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	29/11/2022	29/11/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	29/11/2022	29/11/2022
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	29/11/2022	29/11/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	29/11/2022	29/11/2022
					CT2727	Active implant devices/ancillaries	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	29/11/2022	29/11/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	29/11/2022	29/11/2022
					CT2750	Implantable cardiac electrical stimulators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	29/11/2022	29/11/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	29/11/2022	29/11/2022
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	29/11/2022	29/11/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	29/11/2022	29/11/2022
					CT2985	Active	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	29/11/2022	29/11/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	29/11/2022	29/11/2022
					CT3001	Implantable metal	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	29/11/2022	29/11/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	29/11/2022	29/11/2022
					CT3002	Implantable synthetic polymer	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	29/11/2022	29/11/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	29/11/2022	29/11/2022
					CT3006	Implantable inorganic material	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	29/11/2022	29/11/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	29/11/2022	29/11/2022
					CT314	Electrically powered	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	29/11/2022	29/11/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	29/11/2022	29/11/2022
					CT321	Long-term surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	29/11/2022	29/11/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	29/11/2022	29/11/2022
					CT449	Generators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	29/11/2022	29/11/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	29/11/2022	29/11/2022
					CT981	Single use	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	29/11/2022	29/11/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	29/11/2022	29/11/2022
					CT983	Surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	29/11/2022	29/11/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	29/11/2022	29/11/2022
11858	2022	All other reportable incidents	47265	Dual-chamber implantable pacemaker, rate-responsive	CT141	Electrodes/Leads and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem		
							F1901	Additional Surgery	A24	No Apparent Device Problem		
					CT145	Cardiovascular devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem		
							F1901	Additional Surgery	A24	No Apparent Device Problem		
					CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem		
							F1901	Additional Surgery	A24	No Apparent Device Problem		
					CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem		
							F1901	Additional Surgery	A24	No Apparent Device Problem		
					CT233	Surgical	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem		
							F1901	Additional Surgery	A24	No Apparent Device Problem		
					CT2406	Implantable	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem		
							F1901	Additional Surgery	A24	No Apparent Device Problem		
					CT2407	Active implantable	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem		
							F1901	Additional Surgery	A24	No Apparent Device Problem		
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem		
							F1901	Additional Surgery	A24	No Apparent Device Problem		
					CT2727	Active implant devices/ancillaries	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem		
							F1901	Additional Surgery	A24	No Apparent Device Problem		
					CT2750	Implantable cardiac electrical stimulators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem		
							F1901	Additional Surgery	A24	No Apparent Device Problem		
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem		
							F1901	Additional Surgery	A24	No Apparent Device Problem		
					CT2985	Active	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem		
							F1901	Additional Surgery	A24	No Apparent Device Problem		
					CT3001	Implantable metal	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem		
							F1901	Additional Surgery	A24	No Apparent Device Problem		
					CT3002	Implantable synthetic polymer	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem		
							F1901	Additional Surgery	A24	No Apparent Device Problem		
					CT3006	Implantable inorganic material	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem		
							F1901	Additional Surgery	A24	No Apparent Device Problem		
					CT314	Electrically powered	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem		
							F1901	Additional Surgery	A24	No Apparent Device Problem		
					CT321	Long-term surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem		
							F1901	Additional Surgery	A24	No Apparent Device Problem		
					CT449	Generators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem		
							F1901	Additional Surgery	A24	No Apparent Device Problem		
					CT981	Single use	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem		
							F1901	Additional Surgery	A24	No Apparent Device Problem		
					CT983	Surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem		
							F1901	Additional Surgery	A24	No Apparent Device Problem		
11859	2022	All other reportable incidents	17882	Physiologic-monitoring defibrillation system	CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070803	Failure to Power Up		
							F27	Problem Identified During Non-Clinical Procedure	A070803	Failure to Power Up		
11860	2022		47265	Dual-chamber implantable pacemaker, rate-responsive	CT141	Electrodes/Leads and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	04/12/2022	04/12/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	04/12/2022	04/12/2022
					CT145	Cardiovascular devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	04/12/2022	04/12/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	04/12/2022	04/12/2022
					CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	04/12/2022	04/12/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	04/12/2022	04/12/2022
					CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	04/12/2022	04/12/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	04/12/2022	04/12/2022
					CT233	Surgical	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	04/12/2022	04/12/2022

							F1901	Additional Surgery	A24	No Apparent Device Problem	04/12/2022	04/12/2022
					CT2406	Implantable	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	04/12/2022	04/12/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	04/12/2022	04/12/2022
					CT2407	Active implantable	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	04/12/2022	04/12/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	04/12/2022	04/12/2022
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	04/12/2022	04/12/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	04/12/2022	04/12/2022
					CT2727	Active implant devices/ancillaries	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	04/12/2022	04/12/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	04/12/2022	04/12/2022
					CT2750	Implantable cardiac electrical stimulators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	04/12/2022	04/12/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	04/12/2022	04/12/2022
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	04/12/2022	04/12/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	04/12/2022	04/12/2022
					CT2985	Active	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	04/12/2022	04/12/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	04/12/2022	04/12/2022
					CT3001	Implantable metal	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	04/12/2022	04/12/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	04/12/2022	04/12/2022
					CT3002	Implantable synthetic polymer	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	04/12/2022	04/12/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	04/12/2022	04/12/2022
					CT3006	Implantable inorganic material	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	04/12/2022	04/12/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	04/12/2022	04/12/2022
					CT314	Electrically powered	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	04/12/2022	04/12/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	04/12/2022	04/12/2022
					CT321	Long-term surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	04/12/2022	04/12/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	04/12/2022	04/12/2022
					CT449	Generators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	04/12/2022	04/12/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	04/12/2022	04/12/2022
					CT981	Single use	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	04/12/2022	04/12/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	04/12/2022	04/12/2022
					CT983	Surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	04/12/2022	04/12/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	04/12/2022	04/12/2022
11861	2022	All other reportable incidents	47265	Dual-chamber implantable pacemaker, rate-responsive	CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A0908	Incorrect, Inadequate or Imprecise Result or Readings	25/11/2022	25/11/2022
							F08	Hospitalization or Prolonged Hospitalization	A0908	Incorrect, Inadequate or Imprecise Result or Readings	25/11/2022	25/11/2022
							F1905	Device Revision or Replacement	A0908	Incorrect, Inadequate or Imprecise Result or Readings	25/11/2022	25/11/2022
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A0908	Incorrect, Inadequate or Imprecise Result or Readings	25/11/2022	25/11/2022
							F08	Hospitalization or Prolonged Hospitalization	A0908	Incorrect, Inadequate or Imprecise Result or Readings	25/11/2022	25/11/2022
							F1905	Device Revision or Replacement	A0908	Incorrect, Inadequate or Imprecise Result or Readings	25/11/2022	25/11/2022
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A0908	Incorrect, Inadequate or Imprecise Result or Readings	25/11/2022	25/11/2022
							F08	Hospitalization or Prolonged Hospitalization	A0908	Incorrect, Inadequate or Imprecise Result or Readings	25/11/2022	25/11/2022
							F1905	Device Revision or Replacement	A0908	Incorrect, Inadequate or Imprecise Result or Readings	25/11/2022	25/11/2022
11862	2022	All other reportable incidents	37265	Dual-chamber implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F11	Non-Serious Injury/ Illness/ Impairment	A070504	Premature Discharge of Battery		
11863	2022	All other reportable incidents	17882	Physiologic-monitoring defibrillation system	CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A071301	Failure to Deliver Shock/Stimulation		
							F27	Problem Identified During Non-Clinical Procedure	A071301	Failure to Deliver Shock/Stimulation		
11864	2022	All other reportable incidents	47265	Dual-chamber implantable pacemaker, rate-responsive	CT141	Electrodes/Leads and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	14/12/2022	14/12/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	14/12/2022	14/12/2022
					CT145	Cardiovascular devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	14/12/2022	14/12/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	14/12/2022	14/12/2022
					CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	14/12/2022	14/12/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	14/12/2022	14/12/2022
					CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	14/12/2022	14/12/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	14/12/2022	14/12/2022
					CT233	Surgical	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	14/12/2022	14/12/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	14/12/2022	14/12/2022
					CT2406	Implantable	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	14/12/2022	14/12/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	14/12/2022	14/12/2022
					CT2407	Active implantable	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	14/12/2022	14/12/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	14/12/2022	14/12/2022
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	14/12/2022	14/12/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	14/12/2022	14/12/2022
					CT2727	Active implant devices/ancillaries	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	14/12/2022	14/12/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	14/12/2022	14/12/2022
					CT2750	Implantable cardiac electrical stimulators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	14/12/2022	14/12/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	14/12/2022	14/12/2022
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	14/12/2022	14/12/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	14/12/2022	14/12/2022
					CT2985	Active	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	14/12/2022	14/12/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	14/12/2022	14/12/2022
					CT3001	Implantable metal	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	14/12/2022	14/12/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	14/12/2022	14/12/2022
					CT3002	Implantable synthetic polymer	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	14/12/2022	14/12/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	14/12/2022	14/12/2022
					CT3006	Implantable inorganic material	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	14/12/2022	14/12/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	14/12/2022	14/12/2022
					CT314	Electrically powered	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	14/12/2022	14/12/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	14/12/2022	14/12/2022
					CT321	Long-term surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	14/12/2022	14/12/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	14/12/2022	14/12/2022
					CT449	Generators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	14/12/2022	14/12/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	14/12/2022	14/12/2022
					CT981	Single use	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	14/12/2022	14/12/2022

						F1901	Additional Surgery	A24	No Apparent Device Problem	14/12/2022	14/12/2022
				CT983	Surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	14/12/2022	14/12/2022
						F1901	Additional Surgery	A24	No Apparent Device Problem	14/12/2022	14/12/2022
11865	2022	All other reportable incidents	47265	Dual-chamber implantable pacemaker, rate-responsive	CT141	Electrodes/Leads and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	14/12/2022	14/12/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	14/12/2022	14/12/2022
					CT145	Cardiovascular devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	14/12/2022	14/12/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	14/12/2022	14/12/2022
					CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	14/12/2022	14/12/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	14/12/2022	14/12/2022
					CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	14/12/2022	14/12/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	14/12/2022	14/12/2022
					CT233	Surgical	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	14/12/2022	14/12/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	14/12/2022	14/12/2022
					CT2406	Implantable	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	14/12/2022	14/12/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	14/12/2022	14/12/2022
					CT2407	Active implantable	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	14/12/2022	14/12/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	14/12/2022	14/12/2022
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	14/12/2022	14/12/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	14/12/2022	14/12/2022
					CT2727	Active implant devices/ancillaries	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	14/12/2022	14/12/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	14/12/2022	14/12/2022
					CT2750	Implantable cardiac electrical stimulators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	14/12/2022	14/12/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	14/12/2022	14/12/2022
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	14/12/2022	14/12/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	14/12/2022	14/12/2022
					CT2985	Active	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	14/12/2022	14/12/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	14/12/2022	14/12/2022
					CT3001	Implantable metal	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	14/12/2022	14/12/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	14/12/2022	14/12/2022
					CT3002	Implantable synthetic polymer	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	14/12/2022	14/12/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	14/12/2022	14/12/2022
					CT3006	Implantable inorganic material	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	14/12/2022	14/12/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	14/12/2022	14/12/2022
					CT314	Electrically powered	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	14/12/2022	14/12/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	14/12/2022	14/12/2022
					CT321	Long-term surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	14/12/2022	14/12/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	14/12/2022	14/12/2022
					CT449	Generators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	14/12/2022	14/12/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	14/12/2022	14/12/2022
					CT981	Single use	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	14/12/2022	14/12/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	14/12/2022	14/12/2022
					CT983	Surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	14/12/2022	14/12/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	14/12/2022	14/12/2022
11866	2022	All other reportable incidents	47265	Dual-chamber implantable pacemaker, rate-responsive	CT141	Electrodes/Leads and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	13/12/2022	13/12/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	13/12/2022	13/12/2022
					CT145	Cardiovascular devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	13/12/2022	13/12/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	13/12/2022	13/12/2022
					CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	13/12/2022	13/12/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	13/12/2022	13/12/2022
					CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	13/12/2022	13/12/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	13/12/2022	13/12/2022
					CT233	Surgical	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	13/12/2022	13/12/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	13/12/2022	13/12/2022
					CT2406	Implantable	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	13/12/2022	13/12/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	13/12/2022	13/12/2022
					CT2407	Active implantable	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	13/12/2022	13/12/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	13/12/2022	13/12/2022
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	13/12/2022	13/12/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	13/12/2022	13/12/2022
					CT2727	Active implant devices/ancillaries	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	13/12/2022	13/12/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	13/12/2022	13/12/2022
					CT2750	Implantable cardiac electrical stimulators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	13/12/2022	13/12/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	13/12/2022	13/12/2022
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	13/12/2022	13/12/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	13/12/2022	13/12/2022
					CT2985	Active	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	13/12/2022	13/12/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	13/12/2022	13/12/2022
					CT3001	Implantable metal	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	13/12/2022	13/12/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	13/12/2022	13/12/2022
					CT3002	Implantable synthetic polymer	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	13/12/2022	13/12/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	13/12/2022	13/12/2022
					CT3006	Implantable inorganic material	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	13/12/2022	13/12/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	13/12/2022	13/12/2022
					CT314	Electrically powered	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	13/12/2022	13/12/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	13/12/2022	13/12/2022
					CT321	Long-term surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	13/12/2022	13/12/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	13/12/2022	13/12/2022
					CT449	Generators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	13/12/2022	13/12/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	13/12/2022	13/12/2022
					CT981	Single use	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	13/12/2022	13/12/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	13/12/2022	13/12/2022
					CT983	Surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	13/12/2022	13/12/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	13/12/2022	13/12/2022
11867	2022	All other reportable incidents	35852	Single-chamber implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
11868	2022		47265	Dual-chamber implantable pacemaker, rate-	CT141	Electrodes/Leads and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	10/12/2022	10/12/2022

				responsive			F1901	Additional Surgery	A24	No Apparent Device Problem	10/12/2022	10/12/2022
					CT145	Cardiovascular devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	10/12/2022	10/12/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	10/12/2022	10/12/2022
					CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	10/12/2022	10/12/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	10/12/2022	10/12/2022
					CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	10/12/2022	10/12/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	10/12/2022	10/12/2022
					CT233	Surgical	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	10/12/2022	10/12/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	10/12/2022	10/12/2022
					CT2406	Implantable	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	10/12/2022	10/12/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	10/12/2022	10/12/2022
					CT2407	Active implantable	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	10/12/2022	10/12/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	10/12/2022	10/12/2022
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	10/12/2022	10/12/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	10/12/2022	10/12/2022
					CT2727	Active implant devices/ancillaries	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	10/12/2022	10/12/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	10/12/2022	10/12/2022
					CT2750	Implantable cardiac electrical stimulators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	10/12/2022	10/12/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	10/12/2022	10/12/2022
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	10/12/2022	10/12/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	10/12/2022	10/12/2022
					CT2985	Active	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	10/12/2022	10/12/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	10/12/2022	10/12/2022
					CT3001	Implantable metal	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	10/12/2022	10/12/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	10/12/2022	10/12/2022
					CT3002	Implantable synthetic polymer	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	10/12/2022	10/12/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	10/12/2022	10/12/2022
					CT3006	Implantable inorganic material	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	10/12/2022	10/12/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	10/12/2022	10/12/2022
					CT314	Electrically powered	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	10/12/2022	10/12/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	10/12/2022	10/12/2022
					CT321	Long-term surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	10/12/2022	10/12/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	10/12/2022	10/12/2022
					CT449	Generators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	10/12/2022	10/12/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	10/12/2022	10/12/2022
					CT981	Single use	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	10/12/2022	10/12/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	10/12/2022	10/12/2022
					CT983	Surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	10/12/2022	10/12/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	10/12/2022	10/12/2022
11869	2022	All other reportable incidents	47265	Dual-chamber implantable pacemaker, rate-responsive	CT141	Electrodes/Leads and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	13/12/2022	13/12/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	13/12/2022	13/12/2022
					CT145	Cardiovascular devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	13/12/2022	13/12/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	13/12/2022	13/12/2022
					CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	13/12/2022	13/12/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	13/12/2022	13/12/2022
					CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	13/12/2022	13/12/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	13/12/2022	13/12/2022
					CT233	Surgical	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	13/12/2022	13/12/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	13/12/2022	13/12/2022
					CT2406	Implantable	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	13/12/2022	13/12/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	13/12/2022	13/12/2022
					CT2407	Active implantable	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	13/12/2022	13/12/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	13/12/2022	13/12/2022
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	13/12/2022	13/12/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	13/12/2022	13/12/2022
					CT2727	Active implant devices/ancillaries	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	13/12/2022	13/12/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	13/12/2022	13/12/2022
					CT2750	Implantable cardiac electrical stimulators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	13/12/2022	13/12/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	13/12/2022	13/12/2022
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	13/12/2022	13/12/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	13/12/2022	13/12/2022
					CT2985	Active	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	13/12/2022	13/12/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	13/12/2022	13/12/2022
					CT3001	Implantable metal	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	13/12/2022	13/12/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	13/12/2022	13/12/2022
					CT3002	Implantable synthetic polymer	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	13/12/2022	13/12/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	13/12/2022	13/12/2022
					CT3006	Implantable inorganic material	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	13/12/2022	13/12/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	13/12/2022	13/12/2022
					CT314	Electrically powered	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	13/12/2022	13/12/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	13/12/2022	13/12/2022
					CT321	Long-term surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	13/12/2022	13/12/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	13/12/2022	13/12/2022
					CT449	Generators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	13/12/2022	13/12/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	13/12/2022	13/12/2022
					CT981	Single use	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	13/12/2022	13/12/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	13/12/2022	13/12/2022
					CT983	Surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	13/12/2022	13/12/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	13/12/2022	13/12/2022
11870	2022	All other reportable incidents	47265	Dual-chamber implantable pacemaker, rate-responsive	CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A070101	Failure to Capture	05/12/2022	05/12/2022
									A070102	High Capture Threshold	05/12/2022	05/12/2022
									A0705	Battery Problem	05/12/2022	05/12/2022
									A070504	Premature Discharge of Battery	05/12/2022	05/12/2022
									A070909	Oversensing	05/12/2022	05/12/2022
									A072202	Low Impedance	05/12/2022	05/12/2022
							F08	Hospitalization or Prolonged Hospitalization	A070101	Failure to Capture	05/12/2022	05/12/2022
									A070102	High Capture Threshold	05/12/2022	05/12/2022
									A0705	Battery Problem	05/12/2022	05/12/2022

									A070504	Premature Discharge of Battery	05/12/2022	05/12/2022
									A070909	Oversensing	05/12/2022	05/12/2022
									A072202	Low Impedance	05/12/2022	05/12/2022
							F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A070101	Failure to Capture	05/12/2022	05/12/2022
									A070102	High Capture Threshold	05/12/2022	05/12/2022
									A0705	Battery Problem	05/12/2022	05/12/2022
									A070504	Premature Discharge of Battery	05/12/2022	05/12/2022
									A070909	Oversensing	05/12/2022	05/12/2022
									A072202	Low Impedance	05/12/2022	05/12/2022
							F1905	Device Revision or Replacement	A070101	Failure to Capture	05/12/2022	05/12/2022
									A070102	High Capture Threshold	05/12/2022	05/12/2022
									A0705	Battery Problem	05/12/2022	05/12/2022
									A070504	Premature Discharge of Battery	05/12/2022	05/12/2022
									A070909	Oversensing	05/12/2022	05/12/2022
									A072202	Low Impedance	05/12/2022	05/12/2022
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A070101	Failure to Capture	05/12/2022	05/12/2022
									A070102	High Capture Threshold	05/12/2022	05/12/2022
									A0705	Battery Problem	05/12/2022	05/12/2022
									A070504	Premature Discharge of Battery	05/12/2022	05/12/2022
									A070909	Oversensing	05/12/2022	05/12/2022
									A072202	Low Impedance	05/12/2022	05/12/2022
							F08	Hospitalization or Prolonged Hospitalization	A070101	Failure to Capture	05/12/2022	05/12/2022
									A070102	High Capture Threshold	05/12/2022	05/12/2022
									A0705	Battery Problem	05/12/2022	05/12/2022
									A070504	Premature Discharge of Battery	05/12/2022	05/12/2022
									A070909	Oversensing	05/12/2022	05/12/2022
									A072202	Low Impedance	05/12/2022	05/12/2022
							F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A070101	Failure to Capture	05/12/2022	05/12/2022
									A070102	High Capture Threshold	05/12/2022	05/12/2022
									A0705	Battery Problem	05/12/2022	05/12/2022
									A070504	Premature Discharge of Battery	05/12/2022	05/12/2022
									A070909	Oversensing	05/12/2022	05/12/2022
									A072202	Low Impedance	05/12/2022	05/12/2022
							F1905	Device Revision or Replacement	A070101	Failure to Capture	05/12/2022	05/12/2022
									A070102	High Capture Threshold	05/12/2022	05/12/2022
									A0705	Battery Problem	05/12/2022	05/12/2022
									A070504	Premature Discharge of Battery	05/12/2022	05/12/2022
									A070909	Oversensing	05/12/2022	05/12/2022
									A072202	Low Impedance	05/12/2022	05/12/2022
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A070101	Failure to Capture	05/12/2022	05/12/2022
									A070102	High Capture Threshold	05/12/2022	05/12/2022
									A0705	Battery Problem	05/12/2022	05/12/2022
									A070504	Premature Discharge of Battery	05/12/2022	05/12/2022
									A070909	Oversensing	05/12/2022	05/12/2022
									A072202	Low Impedance	05/12/2022	05/12/2022
							F08	Hospitalization or Prolonged Hospitalization	A070101	Failure to Capture	05/12/2022	05/12/2022
									A070102	High Capture Threshold	05/12/2022	05/12/2022
									A0705	Battery Problem	05/12/2022	05/12/2022
									A070504	Premature Discharge of Battery	05/12/2022	05/12/2022
									A070909	Oversensing	05/12/2022	05/12/2022
									A072202	Low Impedance	05/12/2022	05/12/2022
							F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A070101	Failure to Capture	05/12/2022	05/12/2022
									A070102	High Capture Threshold	05/12/2022	05/12/2022
									A0705	Battery Problem	05/12/2022	05/12/2022
									A070504	Premature Discharge of Battery	05/12/2022	05/12/2022
									A070909	Oversensing	05/12/2022	05/12/2022
									A072202	Low Impedance	05/12/2022	05/12/2022
							F1905	Device Revision or Replacement	A070101	Failure to Capture	05/12/2022	05/12/2022
									A070102	High Capture Threshold	05/12/2022	05/12/2022
									A0705	Battery Problem	05/12/2022	05/12/2022
									A070504	Premature Discharge of Battery	05/12/2022	05/12/2022
									A070909	Oversensing	05/12/2022	05/12/2022
									A072202	Low Impedance	05/12/2022	05/12/2022
11871	2022		47265	Dual-chamber implantable pacemaker, rate-responsive	CT141	Electrodes/Leads and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	28/11/2022	28/11/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	28/11/2022	28/11/2022
					CT145	Cardiovascular devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	28/11/2022	28/11/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	28/11/2022	28/11/2022
					CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	28/11/2022	28/11/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	28/11/2022	28/11/2022
					CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	28/11/2022	28/11/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	28/11/2022	28/11/2022
					CT233	Surgical	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	28/11/2022	28/11/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	28/11/2022	28/11/2022
					CT2406	Implantable	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	28/11/2022	28/11/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	28/11/2022	28/11/2022
					CT2407	Active implantable	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	28/11/2022	28/11/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	28/11/2022	28/11/2022
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	28/11/2022	28/11/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	28/11/2022	28/11/2022
					CT2727	Active implant devices/ancillaries	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	28/11/2022	28/11/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	28/11/2022	28/11/2022
					CT2750	Implantable cardiac electrical stimulators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	28/11/2022	28/11/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	28/11/2022	28/11/2022
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	28/11/2022	28/11/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	28/11/2022	28/11/2022
					CT2985	Active	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	28/11/2022	28/11/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	28/11/2022	28/11/2022
					CT3001	Implantable metal	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	28/11/2022	28/11/2022

						F1901	Additional Surgery	A24	No Apparent Device Problem	28/11/2022	28/11/2022
					CT3002	Implantable synthetic polymer	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	28/11/2022	28/11/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	28/11/2022	28/11/2022
					CT3006	Implantable inorganic material	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	28/11/2022	28/11/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	28/11/2022	28/11/2022
					CT314	Electrically powered	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	28/11/2022	28/11/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	28/11/2022	28/11/2022
					CT321	Long-term surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	28/11/2022	28/11/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	28/11/2022	28/11/2022
					CT449	Generators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	28/11/2022	28/11/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	28/11/2022	28/11/2022
					CT981	Single use	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	28/11/2022	28/11/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	28/11/2022	28/11/2022
					CT983	Surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	28/11/2022	28/11/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	28/11/2022	28/11/2022
11872	2022		47265	Dual-chamber implantable pacemaker, rate-responsive	CT141	Electrodes/Leads and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem		
							F26	No Health Consequences or Impact	A24	No Apparent Device Problem		
					CT145	Cardiovascular devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem		
							F26	No Health Consequences or Impact	A24	No Apparent Device Problem		
					CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem		
							F26	No Health Consequences or Impact	A24	No Apparent Device Problem		
					CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem		
							F26	No Health Consequences or Impact	A24	No Apparent Device Problem		
					CT233	Surgical	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem		
							F26	No Health Consequences or Impact	A24	No Apparent Device Problem		
					CT2406	Implantable	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem		
							F26	No Health Consequences or Impact	A24	No Apparent Device Problem		
					CT2407	Active implantable	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem		
							F26	No Health Consequences or Impact	A24	No Apparent Device Problem		
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem		
							F26	No Health Consequences or Impact	A24	No Apparent Device Problem		
					CT2727	Active implant devices/ancillaries	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem		
							F26	No Health Consequences or Impact	A24	No Apparent Device Problem		
					CT2750	Implantable cardiac electrical stimulators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem		
							F26	No Health Consequences or Impact	A24	No Apparent Device Problem		
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem		
							F26	No Health Consequences or Impact	A24	No Apparent Device Problem		
					CT2985	Active	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem		
							F26	No Health Consequences or Impact	A24	No Apparent Device Problem		
					CT3001	Implantable metal	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem		
							F26	No Health Consequences or Impact	A24	No Apparent Device Problem		
					CT3002	Implantable synthetic polymer	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem		
							F26	No Health Consequences or Impact	A24	No Apparent Device Problem		
					CT3006	Implantable inorganic material	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem		
							F26	No Health Consequences or Impact	A24	No Apparent Device Problem		
					CT314	Electrically powered	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem		
							F26	No Health Consequences or Impact	A24	No Apparent Device Problem		
					CT321	Long-term surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem		
							F26	No Health Consequences or Impact	A24	No Apparent Device Problem		
					CT449	Generators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem		
							F26	No Health Consequences or Impact	A24	No Apparent Device Problem		
					CT981	Single use	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem		
							F26	No Health Consequences or Impact	A24	No Apparent Device Problem		
					CT983	Surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem		
							F26	No Health Consequences or Impact	A24	No Apparent Device Problem		
11873	2022	All other reportable incidents	35223	Endocardial/interventricular septal pacing lead	CT1684	Cardiac pacemakers and associated devices	E0717	Dyspnea	A070101	Failure to Capture		
									A072201	High Impedance		
							F1905	Device Revision or Replacement	A070101	Failure to Capture		
									A072201	High Impedance		
11874	2022		47265	Dual-chamber implantable pacemaker, rate-responsive	CT141	Electrodes/Leads and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	20/12/2022	20/12/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	20/12/2022	20/12/2022
					CT145	Cardiovascular devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	20/12/2022	20/12/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	20/12/2022	20/12/2022
					CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	20/12/2022	20/12/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	20/12/2022	20/12/2022
					CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	20/12/2022	20/12/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	20/12/2022	20/12/2022
					CT233	Surgical	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	20/12/2022	20/12/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	20/12/2022	20/12/2022
					CT2406	Implantable	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	20/12/2022	20/12/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	20/12/2022	20/12/2022
					CT2407	Active implantable	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	20/12/2022	20/12/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	20/12/2022	20/12/2022
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	20/12/2022	20/12/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	20/12/2022	20/12/2022
					CT2727	Active implant devices/ancillaries	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	20/12/2022	20/12/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	20/12/2022	20/12/2022
					CT2750	Implantable cardiac electrical stimulators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	20/12/2022	20/12/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	20/12/2022	20/12/2022
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	20/12/2022	20/12/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	20/12/2022	20/12/2022
					CT2985	Active	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	20/12/2022	20/12/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	20/12/2022	20/12/2022
					CT3001	Implantable metal	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	20/12/2022	20/12/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	20/12/2022	20/12/2022
					CT3002	Implantable synthetic polymer	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	20/12/2022	20/12/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	20/12/2022	20/12/2022
					CT3006	Implantable inorganic material	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	20/12/2022	20/12/2022

						F1901	Additional Surgery	A24	No Apparent Device Problem	20/12/2022	20/12/2022
					CT314	Electrically powered	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	20/12/2022	20/12/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	20/12/2022	20/12/2022
					CT321	Long-term surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	20/12/2022	20/12/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	20/12/2022	20/12/2022
					CT449	Generators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	20/12/2022	20/12/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	20/12/2022	20/12/2022
					CT981	Single use	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	20/12/2022	20/12/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	20/12/2022	20/12/2022
					CT983	Surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	20/12/2022	20/12/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	20/12/2022	20/12/2022
11875	2022	All other reportable incidents	47265	Dual-chamber implantable pacemaker, rate-responsive	CT141	Electrodes/Leads and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	16/12/2022	16/12/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	16/12/2022	16/12/2022
					CT145	Cardiovascular devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	16/12/2022	16/12/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	16/12/2022	16/12/2022
					CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	16/12/2022	16/12/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	16/12/2022	16/12/2022
					CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	16/12/2022	16/12/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	16/12/2022	16/12/2022
					CT233	Surgical	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	16/12/2022	16/12/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	16/12/2022	16/12/2022
					CT2406	Implantable	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	16/12/2022	16/12/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	16/12/2022	16/12/2022
					CT2407	Active implantable	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	16/12/2022	16/12/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	16/12/2022	16/12/2022
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	16/12/2022	16/12/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	16/12/2022	16/12/2022
					CT2727	Active implant devices/ancillaries	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	16/12/2022	16/12/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	16/12/2022	16/12/2022
					CT2750	Implantable cardiac electrical stimulators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	16/12/2022	16/12/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	16/12/2022	16/12/2022
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	16/12/2022	16/12/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	16/12/2022	16/12/2022
					CT2985	Active	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	16/12/2022	16/12/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	16/12/2022	16/12/2022
					CT3001	Implantable metal	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	16/12/2022	16/12/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	16/12/2022	16/12/2022
					CT3002	Implantable synthetic polymer	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	16/12/2022	16/12/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	16/12/2022	16/12/2022
					CT3006	Implantable inorganic material	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	16/12/2022	16/12/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	16/12/2022	16/12/2022
					CT314	Electrically powered	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	16/12/2022	16/12/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	16/12/2022	16/12/2022
					CT321	Long-term surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	16/12/2022	16/12/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	16/12/2022	16/12/2022
					CT449	Generators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	16/12/2022	16/12/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	16/12/2022	16/12/2022
					CT981	Single use	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	16/12/2022	16/12/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	16/12/2022	16/12/2022
					CT983	Surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	16/12/2022	16/12/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	16/12/2022	16/12/2022
11876	2022		47267	Single-chamber implantable pacemaker, rate-responsive	CT141	Electrodes/Leads and associated devices			A0705	Battery Problem		
							E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem		
							F1905	Device Revision or Replacement	A0705	Battery Problem		
					CT145	Cardiovascular devices			A0705	Battery Problem		
							E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem		
							F1905	Device Revision or Replacement	A0705	Battery Problem		
					CT1684	Cardiac pacemakers and associated devices			A0705	Battery Problem		
							E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem		
							F1905	Device Revision or Replacement	A0705	Battery Problem		
					CT1685	Cardiac pacemakers			A0705	Battery Problem		
							E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem		
							F1905	Device Revision or Replacement	A0705	Battery Problem		
					CT233	Surgical			A0705	Battery Problem		
							E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem		
							F1905	Device Revision or Replacement	A0705	Battery Problem		
					CT2406	Implantable			A0705	Battery Problem		
							E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem		
							F1905	Device Revision or Replacement	A0705	Battery Problem		
					CT2407	Active implantable			A0705	Battery Problem		
							E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem		
							F1905	Device Revision or Replacement	A0705	Battery Problem		
					CT257	Implantable cardiac pacemakers			A0705	Battery Problem		
							E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem		
							F1905	Device Revision or Replacement	A0705	Battery Problem		
					CT2727	Active implant devices/ancillaries			A0705	Battery Problem		
							E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem		
							F1905	Device Revision or Replacement	A0705	Battery Problem		
					CT2750	Implantable cardiac electrical stimulators and associated devices			A0705	Battery Problem		
							E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem		
							F1905	Device Revision or Replacement	A0705	Battery Problem		
					CT2866	Implantable cardiac electrical stimulators			A0705	Battery Problem		
							E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem		
							F1905	Device Revision or Replacement	A0705	Battery Problem		
					CT2985	Active			A0705	Battery Problem		
							E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem		
							F1905	Device Revision or Replacement	A0705	Battery Problem		
					CT3001	Implantable metal			A0705	Battery Problem		

						E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem		
						F1905	Device Revision or Replacement	A0705	Battery Problem		
					CT3002	Implantable synthetic polymer			A0705	Battery Problem		
						E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem		
						F1905	Device Revision or Replacement	A0705	Battery Problem		
					CT3006	Implantable inorganic material			A0705	Battery Problem		
						E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem		
						F1905	Device Revision or Replacement	A0705	Battery Problem		
					CT314	Electrically powered			A0705	Battery Problem		
						E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem		
						F1905	Device Revision or Replacement	A0705	Battery Problem		
					CT321	Long-term surgical invasive			A0705	Battery Problem		
						E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem		
						F1905	Device Revision or Replacement	A0705	Battery Problem		
					CT449	Generators and associated devices			A0705	Battery Problem		
						E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem		
						F1905	Device Revision or Replacement	A0705	Battery Problem		
					CT981	Single use			A0705	Battery Problem		
						E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem		
						F1905	Device Revision or Replacement	A0705	Battery Problem		
					CT983	Surgical invasive			A0705	Battery Problem		
						E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem		
						F1905	Device Revision or Replacement	A0705	Battery Problem		
11877	2022	All other reportable incidents	17882	Physiologic-monitoring defibrillation system	CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070803	Failure to Power Up		
							F27	Problem Identified During Non-Clinical Procedure	A070803	Failure to Power Up		
11878	2022	All other reportable incidents	35223	Endocardial/interventricular septal pacing lead	CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A0401	Break		
									A070909	Oversensing		
									A072202	Low Impedance		
							F23	Unexpected Medical Intervention	A0401	Break		
									A070909	Oversensing		
									A072202	Low Impedance		
11879	2022		47265	Dual-chamber implantable pacemaker, rate-responsive	CT141	Electrodes/Leads and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	21/09/2022	21/09/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	21/09/2022	21/09/2022
					CT145	Cardiovascular devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	21/09/2022	21/09/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	21/09/2022	21/09/2022
					CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	21/09/2022	21/09/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	21/09/2022	21/09/2022
					CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	21/09/2022	21/09/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	21/09/2022	21/09/2022
					CT233	Surgical	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	21/09/2022	21/09/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	21/09/2022	21/09/2022
					CT2406	Implantable	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	21/09/2022	21/09/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	21/09/2022	21/09/2022
					CT2407	Active implantable	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	21/09/2022	21/09/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	21/09/2022	21/09/2022
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	21/09/2022	21/09/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	21/09/2022	21/09/2022
					CT2727	Active implant devices/ancillaries	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	21/09/2022	21/09/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	21/09/2022	21/09/2022
					CT2750	Implantable cardiac electrical stimulators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	21/09/2022	21/09/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	21/09/2022	21/09/2022
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	21/09/2022	21/09/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	21/09/2022	21/09/2022
					CT2985	Active	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	21/09/2022	21/09/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	21/09/2022	21/09/2022
					CT3001	Implantable metal	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	21/09/2022	21/09/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	21/09/2022	21/09/2022
					CT3002	Implantable synthetic polymer	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	21/09/2022	21/09/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	21/09/2022	21/09/2022
					CT3006	Implantable inorganic material	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	21/09/2022	21/09/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	21/09/2022	21/09/2022
					CT314	Electrically powered	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	21/09/2022	21/09/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	21/09/2022	21/09/2022
					CT321	Long-term surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	21/09/2022	21/09/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	21/09/2022	21/09/2022
					CT449	Generators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	21/09/2022	21/09/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	21/09/2022	21/09/2022
					CT981	Single use	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	21/09/2022	21/09/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	21/09/2022	21/09/2022
					CT983	Surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	21/09/2022	21/09/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	21/09/2022	21/09/2022
11880	2022	All other reportable incidents	47265	Dual-chamber implantable pacemaker, rate-responsive	CT141	Electrodes/Leads and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	29/11/2022	29/11/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	29/11/2022	29/11/2022
					CT145	Cardiovascular devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	29/11/2022	29/11/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	29/11/2022	29/11/2022
					CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	29/11/2022	29/11/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	29/11/2022	29/11/2022
					CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	29/11/2022	29/11/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	29/11/2022	29/11/2022
					CT233	Surgical	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	29/11/2022	29/11/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	29/11/2022	29/11/2022
					CT2406	Implantable	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	29/11/2022	29/11/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	29/11/2022	29/11/2022
					CT2407	Active implantable	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	29/11/2022	29/11/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	29/11/2022	29/11/2022
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	29/11/2022	29/11/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	29/11/2022	29/11/2022
					CT2727	Active implant devices/ancillaries	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	29/11/2022	29/11/2022

							F1901	Additional Surgery	A24	No Apparent Device Problem	29/11/2022	29/11/2022
					CT2750	Implantable cardiac electrical stimulators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	29/11/2022	29/11/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	29/11/2022	29/11/2022
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	29/11/2022	29/11/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	29/11/2022	29/11/2022
					CT2985	Active	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	29/11/2022	29/11/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	29/11/2022	29/11/2022
					CT3001	Implantable metal	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	29/11/2022	29/11/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	29/11/2022	29/11/2022
					CT3002	Implantable synthetic polymer	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	29/11/2022	29/11/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	29/11/2022	29/11/2022
					CT3006	Implantable inorganic material	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	29/11/2022	29/11/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	29/11/2022	29/11/2022
					CT314	Electrically powered	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	29/11/2022	29/11/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	29/11/2022	29/11/2022
					CT321	Long-term surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	29/11/2022	29/11/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	29/11/2022	29/11/2022
					CT449	Generators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	29/11/2022	29/11/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	29/11/2022	29/11/2022
					CT981	Single use	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	29/11/2022	29/11/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	29/11/2022	29/11/2022
					CT983	Surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	29/11/2022	29/11/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	29/11/2022	29/11/2022
11881	2022	All other reportable incidents	47267	Single-chamber implantable pacemaker, rate-responsive	CT141	Electrodes/Leads and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A071205	Pacemaker Found in Back-Up Mode		
							F1901	Additional Surgery	A071205	Pacemaker Found in Back-Up Mode		
					CT145	Cardiovascular devices	E2403	No Clinical Signs, Symptoms or Conditions	A071205	Pacemaker Found in Back-Up Mode		
							F1901	Additional Surgery	A071205	Pacemaker Found in Back-Up Mode		
					CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A071205	Pacemaker Found in Back-Up Mode		
							F1901	Additional Surgery	A071205	Pacemaker Found in Back-Up Mode		
					CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A071205	Pacemaker Found in Back-Up Mode		
							F1901	Additional Surgery	A071205	Pacemaker Found in Back-Up Mode		
					CT233	Surgical	E2403	No Clinical Signs, Symptoms or Conditions	A071205	Pacemaker Found in Back-Up Mode		
							F1901	Additional Surgery	A071205	Pacemaker Found in Back-Up Mode		
					CT2406	Implantable	E2403	No Clinical Signs, Symptoms or Conditions	A071205	Pacemaker Found in Back-Up Mode		
							F1901	Additional Surgery	A071205	Pacemaker Found in Back-Up Mode		
					CT2407	Active implantable	E2403	No Clinical Signs, Symptoms or Conditions	A071205	Pacemaker Found in Back-Up Mode		
							F1901	Additional Surgery	A071205	Pacemaker Found in Back-Up Mode		
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A071205	Pacemaker Found in Back-Up Mode		
							F1901	Additional Surgery	A071205	Pacemaker Found in Back-Up Mode		
					CT2727	Active implant devices/ancillaries	E2403	No Clinical Signs, Symptoms or Conditions	A071205	Pacemaker Found in Back-Up Mode		
							F1901	Additional Surgery	A071205	Pacemaker Found in Back-Up Mode		
					CT2750	Implantable cardiac electrical stimulators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A071205	Pacemaker Found in Back-Up Mode		
							F1901	Additional Surgery	A071205	Pacemaker Found in Back-Up Mode		
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A071205	Pacemaker Found in Back-Up Mode		
							F1901	Additional Surgery	A071205	Pacemaker Found in Back-Up Mode		
					CT2985	Active	E2403	No Clinical Signs, Symptoms or Conditions	A071205	Pacemaker Found in Back-Up Mode		
							F1901	Additional Surgery	A071205	Pacemaker Found in Back-Up Mode		
					CT3001	Implantable metal	E2403	No Clinical Signs, Symptoms or Conditions	A071205	Pacemaker Found in Back-Up Mode		
							F1901	Additional Surgery	A071205	Pacemaker Found in Back-Up Mode		
					CT3002	Implantable synthetic polymer	E2403	No Clinical Signs, Symptoms or Conditions	A071205	Pacemaker Found in Back-Up Mode		
							F1901	Additional Surgery	A071205	Pacemaker Found in Back-Up Mode		
					CT3006	Implantable inorganic material	E2403	No Clinical Signs, Symptoms or Conditions	A071205	Pacemaker Found in Back-Up Mode		
							F1901	Additional Surgery	A071205	Pacemaker Found in Back-Up Mode		
					CT314	Electrically powered	E2403	No Clinical Signs, Symptoms or Conditions	A071205	Pacemaker Found in Back-Up Mode		
							F1901	Additional Surgery	A071205	Pacemaker Found in Back-Up Mode		
					CT321	Long-term surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A071205	Pacemaker Found in Back-Up Mode		
							F1901	Additional Surgery	A071205	Pacemaker Found in Back-Up Mode		
					CT449	Generators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A071205	Pacemaker Found in Back-Up Mode		
							F1901	Additional Surgery	A071205	Pacemaker Found in Back-Up Mode		
					CT981	Single use	E2403	No Clinical Signs, Symptoms or Conditions	A071205	Pacemaker Found in Back-Up Mode		
							F1901	Additional Surgery	A071205	Pacemaker Found in Back-Up Mode		
					CT983	Surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A071205	Pacemaker Found in Back-Up Mode		
							F1901	Additional Surgery	A071205	Pacemaker Found in Back-Up Mode		
11882	2022	All other reportable incidents	47265	Dual-chamber implantable pacemaker, rate-responsive	CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A090804	False Positive Result		
							F26	No Health Consequences or Impact	A090804	False Positive Result		
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A090804	False Positive Result		
							F26	No Health Consequences or Impact	A090804	False Positive Result		
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A090804	False Positive Result		
							F26	No Health Consequences or Impact	A090804	False Positive Result		
11883	2022	All other reportable incidents	47265	Dual-chamber implantable pacemaker, rate-responsive	CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
11884	2022	All other reportable incidents	47263	Cardiac resynchronization therapy implantable pacemaker	CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A072201	High Impedance		
									A090801	Signal Artifact/Noise		
							F26	No Health Consequences or Impact	A072201	High Impedance		
									A090801	Signal Artifact/Noise		
					CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A072201	High Impedance		
									A090801	Signal Artifact/Noise		
							F26	No Health Consequences or Impact	A072201	High Impedance		
									A090801	Signal Artifact/Noise		
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A072201	High Impedance		
									A090801	Signal Artifact/Noise		
							F26	No Health Consequences or Impact	A072201	High Impedance		
									A090801	Signal Artifact/Noise		
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A072201	High Impedance		

									A090801	Signal Artifact/Noise		
						F26	No Health Consequences or Impact		A072201	High Impedance		
									A090801	Signal Artifact/Noise		
11885	2022		47265	Dual-chamber implantable pacemaker, rate-responsive	CT141	Electrodes/Leads and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	12/12/2022	12/12/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	12/12/2022	12/12/2022
					CT145	Cardiovascular devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	12/12/2022	12/12/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	12/12/2022	12/12/2022
					CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	12/12/2022	12/12/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	12/12/2022	12/12/2022
					CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	12/12/2022	12/12/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	12/12/2022	12/12/2022
					CT233	Surgical	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	12/12/2022	12/12/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	12/12/2022	12/12/2022
					CT2406	Implantable	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	12/12/2022	12/12/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	12/12/2022	12/12/2022
					CT2407	Active implantable	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	12/12/2022	12/12/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	12/12/2022	12/12/2022
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	12/12/2022	12/12/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	12/12/2022	12/12/2022
					CT2727	Active implant devices/ancillaries	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	12/12/2022	12/12/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	12/12/2022	12/12/2022
					CT2750	Implantable cardiac electrical stimulators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	12/12/2022	12/12/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	12/12/2022	12/12/2022
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	12/12/2022	12/12/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	12/12/2022	12/12/2022
					CT2985	Active	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	12/12/2022	12/12/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	12/12/2022	12/12/2022
					CT3001	Implantable metal	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	12/12/2022	12/12/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	12/12/2022	12/12/2022
					CT3002	Implantable synthetic polymer	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	12/12/2022	12/12/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	12/12/2022	12/12/2022
					CT3006	Implantable inorganic material	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	12/12/2022	12/12/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	12/12/2022	12/12/2022
					CT314	Electrically powered	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	12/12/2022	12/12/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	12/12/2022	12/12/2022
					CT321	Long-term surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	12/12/2022	12/12/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	12/12/2022	12/12/2022
					CT449	Generators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	12/12/2022	12/12/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	12/12/2022	12/12/2022
					CT981	Single use	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	12/12/2022	12/12/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	12/12/2022	12/12/2022
					CT983	Surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	12/12/2022	12/12/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	12/12/2022	12/12/2022
11886	2022	All other reportable incidents	47265	Dual-chamber implantable pacemaker, rate-responsive	CT141	Electrodes/Leads and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	13/12/2022	13/12/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	13/12/2022	13/12/2022
					CT145	Cardiovascular devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	13/12/2022	13/12/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	13/12/2022	13/12/2022
					CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	13/12/2022	13/12/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	13/12/2022	13/12/2022
					CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	13/12/2022	13/12/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	13/12/2022	13/12/2022
					CT233	Surgical	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	13/12/2022	13/12/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	13/12/2022	13/12/2022
					CT2406	Implantable	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	13/12/2022	13/12/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	13/12/2022	13/12/2022
					CT2407	Active implantable	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	13/12/2022	13/12/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	13/12/2022	13/12/2022
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	13/12/2022	13/12/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	13/12/2022	13/12/2022
					CT2727	Active implant devices/ancillaries	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	13/12/2022	13/12/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	13/12/2022	13/12/2022
					CT2750	Implantable cardiac electrical stimulators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	13/12/2022	13/12/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	13/12/2022	13/12/2022
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	13/12/2022	13/12/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	13/12/2022	13/12/2022
					CT2985	Active	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	13/12/2022	13/12/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	13/12/2022	13/12/2022
					CT3001	Implantable metal	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	13/12/2022	13/12/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	13/12/2022	13/12/2022
					CT3002	Implantable synthetic polymer	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	13/12/2022	13/12/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	13/12/2022	13/12/2022
					CT3006	Implantable inorganic material	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	13/12/2022	13/12/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	13/12/2022	13/12/2022
					CT314	Electrically powered	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	13/12/2022	13/12/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	13/12/2022	13/12/2022
					CT321	Long-term surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	13/12/2022	13/12/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	13/12/2022	13/12/2022
					CT449	Generators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	13/12/2022	13/12/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	13/12/2022	13/12/2022
					CT981	Single use	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	13/12/2022	13/12/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	13/12/2022	13/12/2022
					CT983	Surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	13/12/2022	13/12/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	13/12/2022	13/12/2022
11887	2022	All other reportable incidents	47265	Dual-chamber implantable pacemaker, rate-responsive	CT141	Electrodes/Leads and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	13/12/2022	13/12/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	13/12/2022	13/12/2022
					CT145	Cardiovascular devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	13/12/2022	13/12/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	13/12/2022	13/12/2022
					CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	13/12/2022	13/12/2022

							F1901	Additional Surgery	A24	No Apparent Device Problem	13/12/2022	13/12/2022
					CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	13/12/2022	13/12/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	13/12/2022	13/12/2022
					CT233	Surgical	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	13/12/2022	13/12/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	13/12/2022	13/12/2022
					CT2406	Implantable	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	13/12/2022	13/12/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	13/12/2022	13/12/2022
					CT2407	Active implantable	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	13/12/2022	13/12/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	13/12/2022	13/12/2022
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	13/12/2022	13/12/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	13/12/2022	13/12/2022
					CT2727	Active implant devices/ancillaries	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	13/12/2022	13/12/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	13/12/2022	13/12/2022
					CT2750	Implantable cardiac electrical stimulators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	13/12/2022	13/12/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	13/12/2022	13/12/2022
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	13/12/2022	13/12/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	13/12/2022	13/12/2022
					CT2985	Active	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	13/12/2022	13/12/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	13/12/2022	13/12/2022
					CT3001	Implantable metal	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	13/12/2022	13/12/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	13/12/2022	13/12/2022
					CT3002	Implantable synthetic polymer	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	13/12/2022	13/12/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	13/12/2022	13/12/2022
					CT3006	Implantable inorganic material	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	13/12/2022	13/12/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	13/12/2022	13/12/2022
					CT314	Electrically powered	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	13/12/2022	13/12/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	13/12/2022	13/12/2022
					CT321	Long-term surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	13/12/2022	13/12/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	13/12/2022	13/12/2022
					CT449	Generators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	13/12/2022	13/12/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	13/12/2022	13/12/2022
					CT981	Single use	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	13/12/2022	13/12/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	13/12/2022	13/12/2022
					CT983	Surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	13/12/2022	13/12/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	13/12/2022	13/12/2022
11888	2022	All other reportable incidents	47265	Dual-chamber implantable pacemaker, rate-responsive	CT141	Electrodes/Leads and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	13/12/2022	13/12/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	13/12/2022	13/12/2022
					CT145	Cardiovascular devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	13/12/2022	13/12/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	13/12/2022	13/12/2022
					CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	13/12/2022	13/12/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	13/12/2022	13/12/2022
					CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	13/12/2022	13/12/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	13/12/2022	13/12/2022
					CT233	Surgical	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	13/12/2022	13/12/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	13/12/2022	13/12/2022
					CT2406	Implantable	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	13/12/2022	13/12/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	13/12/2022	13/12/2022
					CT2407	Active implantable	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	13/12/2022	13/12/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	13/12/2022	13/12/2022
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	13/12/2022	13/12/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	13/12/2022	13/12/2022
					CT2727	Active implant devices/ancillaries	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	13/12/2022	13/12/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	13/12/2022	13/12/2022
					CT2750	Implantable cardiac electrical stimulators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	13/12/2022	13/12/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	13/12/2022	13/12/2022
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	13/12/2022	13/12/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	13/12/2022	13/12/2022
					CT2985	Active	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	13/12/2022	13/12/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	13/12/2022	13/12/2022
					CT3001	Implantable metal	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	13/12/2022	13/12/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	13/12/2022	13/12/2022
					CT3002	Implantable synthetic polymer	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	13/12/2022	13/12/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	13/12/2022	13/12/2022
					CT3006	Implantable inorganic material	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	13/12/2022	13/12/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	13/12/2022	13/12/2022
					CT314	Electrically powered	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	13/12/2022	13/12/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	13/12/2022	13/12/2022
					CT321	Long-term surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	13/12/2022	13/12/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	13/12/2022	13/12/2022
					CT449	Generators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	13/12/2022	13/12/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	13/12/2022	13/12/2022
					CT981	Single use	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	13/12/2022	13/12/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	13/12/2022	13/12/2022
					CT983	Surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	13/12/2022	13/12/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	13/12/2022	13/12/2022
11889	2022	All other reportable incidents	47265	Dual-chamber implantable pacemaker, rate-responsive	CT141	Electrodes/Leads and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	13/12/2022	13/12/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	13/12/2022	13/12/2022
					CT145	Cardiovascular devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	13/12/2022	13/12/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	13/12/2022	13/12/2022
					CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	13/12/2022	13/12/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	13/12/2022	13/12/2022
					CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	13/12/2022	13/12/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	13/12/2022	13/12/2022
					CT233	Surgical	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	13/12/2022	13/12/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	13/12/2022	13/12/2022
					CT2406	Implantable	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	13/12/2022	13/12/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	13/12/2022	13/12/2022
					CT2407	Active implantable	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	13/12/2022	13/12/2022

							F1901	Additional Surgery	A24	No Apparent Device Problem	13/12/2022	13/12/2022
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	13/12/2022	13/12/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	13/12/2022	13/12/2022
					CT2727	Active implant devices/ancillaries	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	13/12/2022	13/12/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	13/12/2022	13/12/2022
					CT2750	Implantable cardiac electrical stimulators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	13/12/2022	13/12/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	13/12/2022	13/12/2022
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	13/12/2022	13/12/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	13/12/2022	13/12/2022
					CT2985	Active	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	13/12/2022	13/12/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	13/12/2022	13/12/2022
					CT3001	Implantable metal	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	13/12/2022	13/12/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	13/12/2022	13/12/2022
					CT3002	Implantable synthetic polymer	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	13/12/2022	13/12/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	13/12/2022	13/12/2022
					CT3006	Implantable inorganic material	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	13/12/2022	13/12/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	13/12/2022	13/12/2022
					CT314	Electrically powered	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	13/12/2022	13/12/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	13/12/2022	13/12/2022
					CT321	Long-term surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	13/12/2022	13/12/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	13/12/2022	13/12/2022
					CT449	Generators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	13/12/2022	13/12/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	13/12/2022	13/12/2022
					CT981	Single use	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	13/12/2022	13/12/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	13/12/2022	13/12/2022
					CT983	Surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	13/12/2022	13/12/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	13/12/2022	13/12/2022
11890	2022	All other reportable incidents	47265	Dual-chamber implantable pacemaker, rate-responsive	CT141	Electrodes/Leads and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	14/12/2022	14/12/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	14/12/2022	14/12/2022
					CT145	Cardiovascular devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	14/12/2022	14/12/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	14/12/2022	14/12/2022
					CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	14/12/2022	14/12/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	14/12/2022	14/12/2022
					CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	14/12/2022	14/12/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	14/12/2022	14/12/2022
					CT233	Surgical	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	14/12/2022	14/12/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	14/12/2022	14/12/2022
					CT2406	Implantable	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	14/12/2022	14/12/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	14/12/2022	14/12/2022
					CT2407	Active implantable	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	14/12/2022	14/12/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	14/12/2022	14/12/2022
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	14/12/2022	14/12/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	14/12/2022	14/12/2022
					CT2727	Active implant devices/ancillaries	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	14/12/2022	14/12/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	14/12/2022	14/12/2022
					CT2750	Implantable cardiac electrical stimulators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	14/12/2022	14/12/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	14/12/2022	14/12/2022
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	14/12/2022	14/12/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	14/12/2022	14/12/2022
					CT2985	Active	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	14/12/2022	14/12/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	14/12/2022	14/12/2022
					CT3001	Implantable metal	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	14/12/2022	14/12/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	14/12/2022	14/12/2022
					CT3002	Implantable synthetic polymer	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	14/12/2022	14/12/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	14/12/2022	14/12/2022
					CT3006	Implantable inorganic material	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	14/12/2022	14/12/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	14/12/2022	14/12/2022
					CT314	Electrically powered	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	14/12/2022	14/12/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	14/12/2022	14/12/2022
					CT321	Long-term surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	14/12/2022	14/12/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	14/12/2022	14/12/2022
					CT449	Generators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	14/12/2022	14/12/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	14/12/2022	14/12/2022
					CT981	Single use	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	14/12/2022	14/12/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	14/12/2022	14/12/2022
					CT983	Surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	14/12/2022	14/12/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	14/12/2022	14/12/2022
11891	2022	All other reportable incidents	17882	Physiologic-monitoring defibrillation system	CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A05	Mechanical Problem		
							F27	Problem Identified During Non-Clinical Procedure	A05	Mechanical Problem		
11892	2022	All other reportable incidents	47265	Dual-chamber implantable pacemaker, rate-responsive	CT141	Electrodes/Leads and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	18/11/2022	18/11/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	18/11/2022	18/11/2022
					CT145	Cardiovascular devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	18/11/2022	18/11/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	18/11/2022	18/11/2022
					CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	18/11/2022	18/11/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	18/11/2022	18/11/2022
					CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	18/11/2022	18/11/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	18/11/2022	18/11/2022
					CT233	Surgical	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	18/11/2022	18/11/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	18/11/2022	18/11/2022
					CT2406	Implantable	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	18/11/2022	18/11/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	18/11/2022	18/11/2022
					CT2407	Active implantable	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	18/11/2022	18/11/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	18/11/2022	18/11/2022
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	18/11/2022	18/11/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	18/11/2022	18/11/2022
					CT2727	Active implant devices/ancillaries	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	18/11/2022	18/11/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	18/11/2022	18/11/2022
					CT2750	Implantable cardiac electrical stimulators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	18/11/2022	18/11/2022

							F1901	Additional Surgery	A24	No Apparent Device Problem	18/11/2022	18/11/2022
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	18/11/2022	18/11/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	18/11/2022	18/11/2022
					CT2985	Active	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	18/11/2022	18/11/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	18/11/2022	18/11/2022
					CT3001	Implantable metal	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	18/11/2022	18/11/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	18/11/2022	18/11/2022
					CT3002	Implantable synthetic polymer	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	18/11/2022	18/11/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	18/11/2022	18/11/2022
					CT3006	Implantable inorganic material	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	18/11/2022	18/11/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	18/11/2022	18/11/2022
					CT314	Electrically powered	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	18/11/2022	18/11/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	18/11/2022	18/11/2022
					CT321	Long-term surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	18/11/2022	18/11/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	18/11/2022	18/11/2022
					CT449	Generators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	18/11/2022	18/11/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	18/11/2022	18/11/2022
					CT981	Single use	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	18/11/2022	18/11/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	18/11/2022	18/11/2022
					CT983	Surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	18/11/2022	18/11/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	18/11/2022	18/11/2022
11893	2022		47265	Dual-chamber implantable pacemaker, rate-responsive	CT141	Electrodes/Leads and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	05/12/2022	05/12/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	05/12/2022	05/12/2022
					CT145	Cardiovascular devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	05/12/2022	05/12/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	05/12/2022	05/12/2022
					CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	05/12/2022	05/12/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	05/12/2022	05/12/2022
					CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	05/12/2022	05/12/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	05/12/2022	05/12/2022
					CT233	Surgical	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	05/12/2022	05/12/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	05/12/2022	05/12/2022
					CT2406	Implantable	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	05/12/2022	05/12/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	05/12/2022	05/12/2022
					CT2407	Active implantable	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	05/12/2022	05/12/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	05/12/2022	05/12/2022
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	05/12/2022	05/12/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	05/12/2022	05/12/2022
					CT2727	Active implant devices/ancillaries	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	05/12/2022	05/12/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	05/12/2022	05/12/2022
					CT2750	Implantable cardiac electrical stimulators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	05/12/2022	05/12/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	05/12/2022	05/12/2022
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	05/12/2022	05/12/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	05/12/2022	05/12/2022
					CT2985	Active	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	05/12/2022	05/12/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	05/12/2022	05/12/2022
					CT3001	Implantable metal	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	05/12/2022	05/12/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	05/12/2022	05/12/2022
					CT3002	Implantable synthetic polymer	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	05/12/2022	05/12/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	05/12/2022	05/12/2022
					CT3006	Implantable inorganic material	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	05/12/2022	05/12/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	05/12/2022	05/12/2022
					CT314	Electrically powered	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	05/12/2022	05/12/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	05/12/2022	05/12/2022
					CT321	Long-term surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	05/12/2022	05/12/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	05/12/2022	05/12/2022
					CT449	Generators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	05/12/2022	05/12/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	05/12/2022	05/12/2022
					CT981	Single use	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	05/12/2022	05/12/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	05/12/2022	05/12/2022
					CT983	Surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	05/12/2022	05/12/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	05/12/2022	05/12/2022
11894	2022		47265	Dual-chamber implantable pacemaker, rate-responsive	CT141	Electrodes/Leads and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	05/12/2022	05/12/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	05/12/2022	05/12/2022
					CT145	Cardiovascular devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	05/12/2022	05/12/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	05/12/2022	05/12/2022
					CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	05/12/2022	05/12/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	05/12/2022	05/12/2022
					CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	05/12/2022	05/12/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	05/12/2022	05/12/2022
					CT233	Surgical	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	05/12/2022	05/12/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	05/12/2022	05/12/2022
					CT2406	Implantable	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	05/12/2022	05/12/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	05/12/2022	05/12/2022
					CT2407	Active implantable	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	05/12/2022	05/12/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	05/12/2022	05/12/2022
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	05/12/2022	05/12/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	05/12/2022	05/12/2022
					CT2727	Active implant devices/ancillaries	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	05/12/2022	05/12/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	05/12/2022	05/12/2022
					CT2750	Implantable cardiac electrical stimulators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	05/12/2022	05/12/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	05/12/2022	05/12/2022
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	05/12/2022	05/12/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	05/12/2022	05/12/2022
					CT2985	Active	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	05/12/2022	05/12/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	05/12/2022	05/12/2022
					CT3001	Implantable metal	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	05/12/2022	05/12/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	05/12/2022	05/12/2022
					CT3002	Implantable synthetic polymer	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	05/12/2022	05/12/2022

						F1901	Additional Surgery	A24	No Apparent Device Problem	05/12/2022	05/12/2022
				CT3006	Implantable inorganic material	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	05/12/2022	05/12/2022
						F1901	Additional Surgery	A24	No Apparent Device Problem	05/12/2022	05/12/2022
				CT314	Electrically powered	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	05/12/2022	05/12/2022
						F1901	Additional Surgery	A24	No Apparent Device Problem	05/12/2022	05/12/2022
				CT321	Long-term surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	05/12/2022	05/12/2022
						F1901	Additional Surgery	A24	No Apparent Device Problem	05/12/2022	05/12/2022
				CT449	Generators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	05/12/2022	05/12/2022
						F1901	Additional Surgery	A24	No Apparent Device Problem	05/12/2022	05/12/2022
				CT981	Single use	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	05/12/2022	05/12/2022
						F1901	Additional Surgery	A24	No Apparent Device Problem	05/12/2022	05/12/2022
				CT983	Surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	05/12/2022	05/12/2022
						F1901	Additional Surgery	A24	No Apparent Device Problem	05/12/2022	05/12/2022
11895	2022	All other reportable incidents	47265	Dual-chamber implantable pacemaker, rate-responsive	CT141	Electrodes/Leads and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem		
							F1901	Additional Surgery	A24	No Apparent Device Problem		
					CT145	Cardiovascular devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem		
							F1901	Additional Surgery	A24	No Apparent Device Problem		
					CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem		
							F1901	Additional Surgery	A24	No Apparent Device Problem		
					CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem		
							F1901	Additional Surgery	A24	No Apparent Device Problem		
					CT233	Surgical	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem		
							F1901	Additional Surgery	A24	No Apparent Device Problem		
					CT2406	Implantable	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem		
							F1901	Additional Surgery	A24	No Apparent Device Problem		
					CT2407	Active implantable	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem		
							F1901	Additional Surgery	A24	No Apparent Device Problem		
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem		
							F1901	Additional Surgery	A24	No Apparent Device Problem		
					CT2727	Active implant devices/ancillaries	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem		
							F1901	Additional Surgery	A24	No Apparent Device Problem		
					CT2750	Implantable cardiac electrical stimulators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem		
							F1901	Additional Surgery	A24	No Apparent Device Problem		
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem		
							F1901	Additional Surgery	A24	No Apparent Device Problem		
					CT2985	Active	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem		
							F1901	Additional Surgery	A24	No Apparent Device Problem		
					CT3001	Implantable metal	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem		
							F1901	Additional Surgery	A24	No Apparent Device Problem		
					CT3002	Implantable synthetic polymer	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem		
							F1901	Additional Surgery	A24	No Apparent Device Problem		
					CT3006	Implantable inorganic material	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem		
							F1901	Additional Surgery	A24	No Apparent Device Problem		
					CT314	Electrically powered	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem		
							F1901	Additional Surgery	A24	No Apparent Device Problem		
					CT321	Long-term surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem		
							F1901	Additional Surgery	A24	No Apparent Device Problem		
					CT449	Generators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem		
							F1901	Additional Surgery	A24	No Apparent Device Problem		
					CT981	Single use	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem		
							F1901	Additional Surgery	A24	No Apparent Device Problem		
					CT983	Surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem		
							F1901	Additional Surgery	A24	No Apparent Device Problem		
11896	2022		47265	Dual-chamber implantable pacemaker, rate-responsive	CT141	Electrodes/Leads and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	10/11/2022	10/11/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	10/11/2022	10/11/2022
					CT145	Cardiovascular devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	10/11/2022	10/11/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	10/11/2022	10/11/2022
					CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	10/11/2022	10/11/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	10/11/2022	10/11/2022
					CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	10/11/2022	10/11/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	10/11/2022	10/11/2022
					CT233	Surgical	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	10/11/2022	10/11/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	10/11/2022	10/11/2022
					CT2406	Implantable	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	10/11/2022	10/11/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	10/11/2022	10/11/2022
					CT2407	Active implantable	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	10/11/2022	10/11/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	10/11/2022	10/11/2022
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	10/11/2022	10/11/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	10/11/2022	10/11/2022
					CT2727	Active implant devices/ancillaries	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	10/11/2022	10/11/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	10/11/2022	10/11/2022
					CT2750	Implantable cardiac electrical stimulators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	10/11/2022	10/11/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	10/11/2022	10/11/2022
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	10/11/2022	10/11/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	10/11/2022	10/11/2022
					CT2985	Active	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	10/11/2022	10/11/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	10/11/2022	10/11/2022
					CT3001	Implantable metal	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	10/11/2022	10/11/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	10/11/2022	10/11/2022
					CT3002	Implantable synthetic polymer	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	10/11/2022	10/11/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	10/11/2022	10/11/2022
					CT3006	Implantable inorganic material	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	10/11/2022	10/11/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	10/11/2022	10/11/2022
					CT314	Electrically powered	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	10/11/2022	10/11/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	10/11/2022	10/11/2022
					CT321	Long-term surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	10/11/2022	10/11/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	10/11/2022	10/11/2022
					CT449	Generators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	10/11/2022	10/11/2022

						F1901	Additional Surgery	A24	No Apparent Device Problem	10/11/2022	10/11/2022
				CT981	Single use	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	10/11/2022	10/11/2022
						F1901	Additional Surgery	A24	No Apparent Device Problem	10/11/2022	10/11/2022
				CT983	Surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	10/11/2022	10/11/2022
						F1901	Additional Surgery	A24	No Apparent Device Problem	10/11/2022	10/11/2022
11897	2022		47265	Dual-chamber implantable pacemaker, rate-responsive	CT141	Electrodes/Leads and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	31/10/2022	31/10/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	31/10/2022	31/10/2022
					CT145	Cardiovascular devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	31/10/2022	31/10/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	31/10/2022	31/10/2022
					CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	31/10/2022	31/10/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	31/10/2022	31/10/2022
					CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	31/10/2022	31/10/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	31/10/2022	31/10/2022
					CT233	Surgical	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	31/10/2022	31/10/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	31/10/2022	31/10/2022
					CT2406	Implantable	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	31/10/2022	31/10/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	31/10/2022	31/10/2022
					CT2407	Active implantable	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	31/10/2022	31/10/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	31/10/2022	31/10/2022
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	31/10/2022	31/10/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	31/10/2022	31/10/2022
					CT2727	Active implant devices/ancillaries	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	31/10/2022	31/10/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	31/10/2022	31/10/2022
					CT2750	Implantable cardiac electrical stimulators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	31/10/2022	31/10/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	31/10/2022	31/10/2022
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	31/10/2022	31/10/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	31/10/2022	31/10/2022
					CT2985	Active	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	31/10/2022	31/10/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	31/10/2022	31/10/2022
					CT3001	Implantable metal	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	31/10/2022	31/10/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	31/10/2022	31/10/2022
					CT3002	Implantable synthetic polymer	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	31/10/2022	31/10/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	31/10/2022	31/10/2022
					CT3006	Implantable inorganic material	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	31/10/2022	31/10/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	31/10/2022	31/10/2022
					CT314	Electrically powered	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	31/10/2022	31/10/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	31/10/2022	31/10/2022
					CT321	Long-term surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	31/10/2022	31/10/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	31/10/2022	31/10/2022
					CT449	Generators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	31/10/2022	31/10/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	31/10/2022	31/10/2022
					CT981	Single use	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	31/10/2022	31/10/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	31/10/2022	31/10/2022
					CT983	Surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	31/10/2022	31/10/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	31/10/2022	31/10/2022
11898	2022		47265	Dual-chamber implantable pacemaker, rate-responsive	CT141	Electrodes/Leads and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	30/11/2022	30/11/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	30/11/2022	30/11/2022
					CT145	Cardiovascular devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	30/11/2022	30/11/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	30/11/2022	30/11/2022
					CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	30/11/2022	30/11/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	30/11/2022	30/11/2022
					CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	30/11/2022	30/11/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	30/11/2022	30/11/2022
					CT233	Surgical	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	30/11/2022	30/11/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	30/11/2022	30/11/2022
					CT2406	Implantable	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	30/11/2022	30/11/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	30/11/2022	30/11/2022
					CT2407	Active implantable	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	30/11/2022	30/11/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	30/11/2022	30/11/2022
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	30/11/2022	30/11/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	30/11/2022	30/11/2022
					CT2727	Active implant devices/ancillaries	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	30/11/2022	30/11/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	30/11/2022	30/11/2022
					CT2750	Implantable cardiac electrical stimulators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	30/11/2022	30/11/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	30/11/2022	30/11/2022
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	30/11/2022	30/11/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	30/11/2022	30/11/2022
					CT2985	Active	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	30/11/2022	30/11/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	30/11/2022	30/11/2022
					CT3001	Implantable metal	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	30/11/2022	30/11/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	30/11/2022	30/11/2022
					CT3002	Implantable synthetic polymer	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	30/11/2022	30/11/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	30/11/2022	30/11/2022
					CT3006	Implantable inorganic material	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	30/11/2022	30/11/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	30/11/2022	30/11/2022
					CT314	Electrically powered	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	30/11/2022	30/11/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	30/11/2022	30/11/2022
					CT321	Long-term surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	30/11/2022	30/11/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	30/11/2022	30/11/2022
					CT449	Generators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	30/11/2022	30/11/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	30/11/2022	30/11/2022
					CT981	Single use	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	30/11/2022	30/11/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	30/11/2022	30/11/2022
					CT983	Surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	30/11/2022	30/11/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	30/11/2022	30/11/2022
11899	2022		47265	Dual-chamber implantable pacemaker, rate-responsive	CT141	Electrodes/Leads and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	05/12/2022	05/12/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	05/12/2022	05/12/2022
					CT145	Cardiovascular devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	05/12/2022	05/12/2022

							F1901	Additional Surgery	A24	No Apparent Device Problem	05/12/2022	05/12/2022
					CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	05/12/2022	05/12/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	05/12/2022	05/12/2022
					CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	05/12/2022	05/12/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	05/12/2022	05/12/2022
					CT233	Surgical	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	05/12/2022	05/12/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	05/12/2022	05/12/2022
					CT2406	Implantable	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	05/12/2022	05/12/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	05/12/2022	05/12/2022
					CT2407	Active implantable	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	05/12/2022	05/12/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	05/12/2022	05/12/2022
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	05/12/2022	05/12/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	05/12/2022	05/12/2022
					CT2727	Active implant devices/ancillaries	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	05/12/2022	05/12/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	05/12/2022	05/12/2022
					CT2750	Implantable cardiac electrical stimulators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	05/12/2022	05/12/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	05/12/2022	05/12/2022
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	05/12/2022	05/12/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	05/12/2022	05/12/2022
					CT2985	Active	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	05/12/2022	05/12/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	05/12/2022	05/12/2022
					CT3001	Implantable metal	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	05/12/2022	05/12/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	05/12/2022	05/12/2022
					CT3002	Implantable synthetic polymer	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	05/12/2022	05/12/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	05/12/2022	05/12/2022
					CT3006	Implantable inorganic material	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	05/12/2022	05/12/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	05/12/2022	05/12/2022
					CT314	Electrically powered	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	05/12/2022	05/12/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	05/12/2022	05/12/2022
					CT321	Long-term surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	05/12/2022	05/12/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	05/12/2022	05/12/2022
					CT449	Generators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	05/12/2022	05/12/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	05/12/2022	05/12/2022
					CT981	Single use	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	05/12/2022	05/12/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	05/12/2022	05/12/2022
					CT983	Surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	05/12/2022	05/12/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	05/12/2022	05/12/2022
11900	2022		47265	Dual-chamber implantable pacemaker, rate-responsive	CT141	Electrodes/Leads and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	05/12/2022	05/12/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	05/12/2022	05/12/2022
					CT145	Cardiovascular devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	05/12/2022	05/12/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	05/12/2022	05/12/2022
					CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	05/12/2022	05/12/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	05/12/2022	05/12/2022
					CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	05/12/2022	05/12/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	05/12/2022	05/12/2022
					CT233	Surgical	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	05/12/2022	05/12/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	05/12/2022	05/12/2022
					CT2406	Implantable	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	05/12/2022	05/12/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	05/12/2022	05/12/2022
					CT2407	Active implantable	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	05/12/2022	05/12/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	05/12/2022	05/12/2022
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	05/12/2022	05/12/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	05/12/2022	05/12/2022
					CT2727	Active implant devices/ancillaries	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	05/12/2022	05/12/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	05/12/2022	05/12/2022
					CT2750	Implantable cardiac electrical stimulators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	05/12/2022	05/12/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	05/12/2022	05/12/2022
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	05/12/2022	05/12/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	05/12/2022	05/12/2022
					CT2985	Active	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	05/12/2022	05/12/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	05/12/2022	05/12/2022
					CT3001	Implantable metal	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	05/12/2022	05/12/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	05/12/2022	05/12/2022
					CT3002	Implantable synthetic polymer	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	05/12/2022	05/12/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	05/12/2022	05/12/2022
					CT3006	Implantable inorganic material	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	05/12/2022	05/12/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	05/12/2022	05/12/2022
					CT314	Electrically powered	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	05/12/2022	05/12/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	05/12/2022	05/12/2022
					CT321	Long-term surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	05/12/2022	05/12/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	05/12/2022	05/12/2022
					CT449	Generators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	05/12/2022	05/12/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	05/12/2022	05/12/2022
					CT981	Single use	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	05/12/2022	05/12/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	05/12/2022	05/12/2022
					CT983	Surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	05/12/2022	05/12/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	05/12/2022	05/12/2022
11901	2022		47265	Dual-chamber implantable pacemaker, rate-responsive	CT141	Electrodes/Leads and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	02/12/2022	02/12/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	02/12/2022	02/12/2022
					CT145	Cardiovascular devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	02/12/2022	02/12/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	02/12/2022	02/12/2022
					CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	02/12/2022	02/12/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	02/12/2022	02/12/2022
					CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	02/12/2022	02/12/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	02/12/2022	02/12/2022
					CT233	Surgical	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	02/12/2022	02/12/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	02/12/2022	02/12/2022
					CT2406	Implantable	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	02/12/2022	02/12/2022

							F1901	Additional Surgery	A24	No Apparent Device Problem	02/12/2022	02/12/2022
					CT2407	Active implantable	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	02/12/2022	02/12/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	02/12/2022	02/12/2022
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	02/12/2022	02/12/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	02/12/2022	02/12/2022
					CT2727	Active implant devices/ancillaries	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	02/12/2022	02/12/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	02/12/2022	02/12/2022
					CT2750	Implantable cardiac electrical stimulators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	02/12/2022	02/12/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	02/12/2022	02/12/2022
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	02/12/2022	02/12/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	02/12/2022	02/12/2022
					CT2985	Active	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	02/12/2022	02/12/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	02/12/2022	02/12/2022
					CT3001	Implantable metal	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	02/12/2022	02/12/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	02/12/2022	02/12/2022
					CT3002	Implantable synthetic polymer	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	02/12/2022	02/12/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	02/12/2022	02/12/2022
					CT3006	Implantable inorganic material	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	02/12/2022	02/12/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	02/12/2022	02/12/2022
					CT314	Electrically powered	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	02/12/2022	02/12/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	02/12/2022	02/12/2022
					CT321	Long-term surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	02/12/2022	02/12/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	02/12/2022	02/12/2022
					CT449	Generators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	02/12/2022	02/12/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	02/12/2022	02/12/2022
					CT981	Single use	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	02/12/2022	02/12/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	02/12/2022	02/12/2022
					CT983	Surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	02/12/2022	02/12/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	02/12/2022	02/12/2022
11902	2022		47265	Dual-chamber implantable pacemaker, rate-responsive	CT141	Electrodes/Leads and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	02/12/2022	02/12/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	02/12/2022	02/12/2022
					CT145	Cardiovascular devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	02/12/2022	02/12/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	02/12/2022	02/12/2022
					CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	02/12/2022	02/12/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	02/12/2022	02/12/2022
					CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	02/12/2022	02/12/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	02/12/2022	02/12/2022
					CT233	Surgical	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	02/12/2022	02/12/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	02/12/2022	02/12/2022
					CT2406	Implantable	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	02/12/2022	02/12/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	02/12/2022	02/12/2022
					CT2407	Active implantable	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	02/12/2022	02/12/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	02/12/2022	02/12/2022
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	02/12/2022	02/12/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	02/12/2022	02/12/2022
					CT2727	Active implant devices/ancillaries	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	02/12/2022	02/12/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	02/12/2022	02/12/2022
					CT2750	Implantable cardiac electrical stimulators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	02/12/2022	02/12/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	02/12/2022	02/12/2022
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	02/12/2022	02/12/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	02/12/2022	02/12/2022
					CT2985	Active	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	02/12/2022	02/12/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	02/12/2022	02/12/2022
					CT3001	Implantable metal	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	02/12/2022	02/12/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	02/12/2022	02/12/2022
					CT3002	Implantable synthetic polymer	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	02/12/2022	02/12/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	02/12/2022	02/12/2022
					CT3006	Implantable inorganic material	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	02/12/2022	02/12/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	02/12/2022	02/12/2022
					CT314	Electrically powered	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	02/12/2022	02/12/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	02/12/2022	02/12/2022
					CT321	Long-term surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	02/12/2022	02/12/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	02/12/2022	02/12/2022
					CT449	Generators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	02/12/2022	02/12/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	02/12/2022	02/12/2022
					CT981	Single use	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	02/12/2022	02/12/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	02/12/2022	02/12/2022
					CT983	Surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	02/12/2022	02/12/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	02/12/2022	02/12/2022
11903	2022	All other reportable incidents	17882	Physiologic-monitoring defibrillation system	CT1684	Cardiac pacemakers and associated devices	E060110	Ventricular Fibrillation	A0713	Defibrillation/Stimulation Problem		
							F26	No Health Consequences or Impact	A0713	Defibrillation/Stimulation Problem		
11904	2022	All other reportable incidents	47265	Dual-chamber implantable pacemaker, rate-responsive	CT141	Electrodes/Leads and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	02/12/2022	02/12/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	02/12/2022	02/12/2022
					CT145	Cardiovascular devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	02/12/2022	02/12/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	02/12/2022	02/12/2022
					CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	02/12/2022	02/12/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	02/12/2022	02/12/2022
					CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	02/12/2022	02/12/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	02/12/2022	02/12/2022
					CT233	Surgical	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	02/12/2022	02/12/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	02/12/2022	02/12/2022
					CT2406	Implantable	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	02/12/2022	02/12/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	02/12/2022	02/12/2022
					CT2407	Active implantable	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	02/12/2022	02/12/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	02/12/2022	02/12/2022
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	02/12/2022	02/12/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	02/12/2022	02/12/2022
					CT2727	Active implant devices/ancillaries	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	02/12/2022	02/12/2022

							F1901	Additional Surgery	A24	No Apparent Device Problem	02/12/2022	02/12/2022
					CT2750	Implantable cardiac electrical stimulators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	02/12/2022	02/12/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	02/12/2022	02/12/2022
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	02/12/2022	02/12/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	02/12/2022	02/12/2022
					CT2985	Active	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	02/12/2022	02/12/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	02/12/2022	02/12/2022
					CT3001	Implantable metal	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	02/12/2022	02/12/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	02/12/2022	02/12/2022
					CT3002	Implantable synthetic polymer	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	02/12/2022	02/12/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	02/12/2022	02/12/2022
					CT3006	Implantable inorganic material	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	02/12/2022	02/12/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	02/12/2022	02/12/2022
					CT314	Electrically powered	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	02/12/2022	02/12/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	02/12/2022	02/12/2022
					CT321	Long-term surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	02/12/2022	02/12/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	02/12/2022	02/12/2022
					CT449	Generators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	02/12/2022	02/12/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	02/12/2022	02/12/2022
					CT981	Single use	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	02/12/2022	02/12/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	02/12/2022	02/12/2022
					CT983	Surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	02/12/2022	02/12/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	02/12/2022	02/12/2022
11905	2022		47265	Dual-chamber implantable pacemaker, rate-responsive	CT141	Electrodes/Leads and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	02/12/2022	02/12/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	02/12/2022	02/12/2022
					CT145	Cardiovascular devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	02/12/2022	02/12/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	02/12/2022	02/12/2022
					CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	02/12/2022	02/12/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	02/12/2022	02/12/2022
					CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	02/12/2022	02/12/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	02/12/2022	02/12/2022
					CT233	Surgical	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	02/12/2022	02/12/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	02/12/2022	02/12/2022
					CT2406	Implantable	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	02/12/2022	02/12/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	02/12/2022	02/12/2022
					CT2407	Active implantable	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	02/12/2022	02/12/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	02/12/2022	02/12/2022
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	02/12/2022	02/12/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	02/12/2022	02/12/2022
					CT2727	Active implant devices/ancillaries	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	02/12/2022	02/12/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	02/12/2022	02/12/2022
					CT2750	Implantable cardiac electrical stimulators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	02/12/2022	02/12/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	02/12/2022	02/12/2022
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	02/12/2022	02/12/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	02/12/2022	02/12/2022
					CT2985	Active	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	02/12/2022	02/12/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	02/12/2022	02/12/2022
					CT3001	Implantable metal	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	02/12/2022	02/12/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	02/12/2022	02/12/2022
					CT3002	Implantable synthetic polymer	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	02/12/2022	02/12/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	02/12/2022	02/12/2022
					CT3006	Implantable inorganic material	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	02/12/2022	02/12/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	02/12/2022	02/12/2022
					CT314	Electrically powered	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	02/12/2022	02/12/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	02/12/2022	02/12/2022
					CT321	Long-term surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	02/12/2022	02/12/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	02/12/2022	02/12/2022
					CT449	Generators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	02/12/2022	02/12/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	02/12/2022	02/12/2022
					CT981	Single use	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	02/12/2022	02/12/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	02/12/2022	02/12/2022
					CT983	Surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	02/12/2022	02/12/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	02/12/2022	02/12/2022
11906	2022	All other reportable incidents	47265	Dual-chamber implantable pacemaker, rate-responsive	CT141	Electrodes/Leads and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	06/12/2022	06/12/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	06/12/2022	06/12/2022
					CT145	Cardiovascular devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	06/12/2022	06/12/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	06/12/2022	06/12/2022
					CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	06/12/2022	06/12/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	06/12/2022	06/12/2022
					CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	06/12/2022	06/12/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	06/12/2022	06/12/2022
					CT233	Surgical	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	06/12/2022	06/12/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	06/12/2022	06/12/2022
					CT2406	Implantable	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	06/12/2022	06/12/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	06/12/2022	06/12/2022
					CT2407	Active implantable	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	06/12/2022	06/12/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	06/12/2022	06/12/2022
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	06/12/2022	06/12/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	06/12/2022	06/12/2022
					CT2727	Active implant devices/ancillaries	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	06/12/2022	06/12/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	06/12/2022	06/12/2022
					CT2750	Implantable cardiac electrical stimulators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	06/12/2022	06/12/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	06/12/2022	06/12/2022
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	06/12/2022	06/12/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	06/12/2022	06/12/2022
					CT2985	Active	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	06/12/2022	06/12/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	06/12/2022	06/12/2022
					CT3001	Implantable metal	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	06/12/2022	06/12/2022

						F1901	Additional Surgery	A24	No Apparent Device Problem	06/12/2022	06/12/2022
				CT3002	Implantable synthetic polymer	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	06/12/2022	06/12/2022
						F1901	Additional Surgery	A24	No Apparent Device Problem	06/12/2022	06/12/2022
				CT3006	Implantable inorganic material	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	06/12/2022	06/12/2022
						F1901	Additional Surgery	A24	No Apparent Device Problem	06/12/2022	06/12/2022
				CT314	Electrically powered	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	06/12/2022	06/12/2022
						F1901	Additional Surgery	A24	No Apparent Device Problem	06/12/2022	06/12/2022
				CT321	Long-term surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	06/12/2022	06/12/2022
						F1901	Additional Surgery	A24	No Apparent Device Problem	06/12/2022	06/12/2022
				CT449	Generators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	06/12/2022	06/12/2022
						F1901	Additional Surgery	A24	No Apparent Device Problem	06/12/2022	06/12/2022
				CT981	Single use	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	06/12/2022	06/12/2022
						F1901	Additional Surgery	A24	No Apparent Device Problem	06/12/2022	06/12/2022
				CT983	Surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	06/12/2022	06/12/2022
						F1901	Additional Surgery	A24	No Apparent Device Problem	06/12/2022	06/12/2022
11907	2022	All other reportable incidents	17882	Physiologic-monitoring defibrillation system	CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A071204	No Pacing		
							F27	Problem Identified During Non-Clinical Procedure	A071204	No Pacing		
11908	2022	All other reportable incidents	47265	Dual-chamber implantable pacemaker, rate-responsive	CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem	15/09/2022	15/09/2022
									A070504	Premature Discharge of Battery	15/09/2022	15/09/2022
							F08	Hospitalization or Prolonged Hospitalization	A0705	Battery Problem	15/09/2022	15/09/2022
									A070504	Premature Discharge of Battery	15/09/2022	15/09/2022
							F1905	Device Revision or Replacement	A0705	Battery Problem	15/09/2022	15/09/2022
									A070504	Premature Discharge of Battery	15/09/2022	15/09/2022
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem	15/09/2022	15/09/2022
									A070504	Premature Discharge of Battery	15/09/2022	15/09/2022
							F08	Hospitalization or Prolonged Hospitalization	A0705	Battery Problem	15/09/2022	15/09/2022
									A070504	Premature Discharge of Battery	15/09/2022	15/09/2022
							F1905	Device Revision or Replacement	A0705	Battery Problem	15/09/2022	15/09/2022
									A070504	Premature Discharge of Battery	15/09/2022	15/09/2022
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem	15/09/2022	15/09/2022
									A070504	Premature Discharge of Battery	15/09/2022	15/09/2022
							F08	Hospitalization or Prolonged Hospitalization	A0705	Battery Problem	15/09/2022	15/09/2022
									A070504	Premature Discharge of Battery	15/09/2022	15/09/2022
							F1905	Device Revision or Replacement	A0705	Battery Problem	15/09/2022	15/09/2022
									A070504	Premature Discharge of Battery	15/09/2022	15/09/2022
11909	2022		47264	Dual-chamber implantable pacemaker, demand	CT141	Electrodes/Leads and associated devices	F11	Non-Serious Injury/ Illness/ Impairment	A0705	Battery Problem		
					CT145	Cardiovascular devices	F11	Non-Serious Injury/ Illness/ Impairment	A0705	Battery Problem		
					CT1684	Cardiac pacemakers and associated devices	F11	Non-Serious Injury/ Illness/ Impairment	A0705	Battery Problem		
					CT1685	Cardiac pacemakers	F11	Non-Serious Injury/ Illness/ Impairment	A0705	Battery Problem		
					CT233	Surgical	F11	Non-Serious Injury/ Illness/ Impairment	A0705	Battery Problem		
					CT2406	Implantable	F11	Non-Serious Injury/ Illness/ Impairment	A0705	Battery Problem		
					CT2407	Active implantable	F11	Non-Serious Injury/ Illness/ Impairment	A0705	Battery Problem		
					CT257	Implantable cardiac pacemakers	F11	Non-Serious Injury/ Illness/ Impairment	A0705	Battery Problem		
					CT2727	Active implant devices/ancillaries	F11	Non-Serious Injury/ Illness/ Impairment	A0705	Battery Problem		
					CT2750	Implantable cardiac electrical stimulators and associated devices	F11	Non-Serious Injury/ Illness/ Impairment	A0705	Battery Problem		
					CT2866	Implantable cardiac electrical stimulators	F11	Non-Serious Injury/ Illness/ Impairment	A0705	Battery Problem		
					CT2985	Active	F11	Non-Serious Injury/ Illness/ Impairment	A0705	Battery Problem		
					CT3001	Implantable metal	F11	Non-Serious Injury/ Illness/ Impairment	A0705	Battery Problem		
					CT3002	Implantable synthetic polymer	F11	Non-Serious Injury/ Illness/ Impairment	A0705	Battery Problem		
					CT3006	Implantable inorganic material	F11	Non-Serious Injury/ Illness/ Impairment	A0705	Battery Problem		
					CT314	Electrically powered	F11	Non-Serious Injury/ Illness/ Impairment	A0705	Battery Problem		
					CT321	Long-term surgical invasive	F11	Non-Serious Injury/ Illness/ Impairment	A0705	Battery Problem		
					CT449	Generators and associated devices	F11	Non-Serious Injury/ Illness/ Impairment	A0705	Battery Problem		
					CT981	Single use	F11	Non-Serious Injury/ Illness/ Impairment	A0705	Battery Problem		
					CT983	Surgical invasive	F11	Non-Serious Injury/ Illness/ Impairment	A0705	Battery Problem		
11910	2022		47265	Dual-chamber implantable pacemaker, rate-responsive	CT141	Electrodes/Leads and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem		
							F1901	Additional Surgery	A24	No Apparent Device Problem		
					CT145	Cardiovascular devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem		
							F1901	Additional Surgery	A24	No Apparent Device Problem		
					CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem		
							F1901	Additional Surgery	A24	No Apparent Device Problem		
					CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem		
							F1901	Additional Surgery	A24	No Apparent Device Problem		
					CT233	Surgical	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem		
							F1901	Additional Surgery	A24	No Apparent Device Problem		
					CT2406	Implantable	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem		
							F1901	Additional Surgery	A24	No Apparent Device Problem		
					CT2407	Active implantable	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem		
							F1901	Additional Surgery	A24	No Apparent Device Problem		
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem		
							F1901	Additional Surgery	A24	No Apparent Device Problem		
					CT2727	Active implant devices/ancillaries	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem		
							F1901	Additional Surgery	A24	No Apparent Device Problem		
					CT2750	Implantable cardiac electrical stimulators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem		
							F1901	Additional Surgery	A24	No Apparent Device Problem		
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem		
							F1901	Additional Surgery	A24	No Apparent Device Problem		
					CT2985	Active	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem		
							F1901	Additional Surgery	A24	No Apparent Device Problem		
					CT3001	Implantable metal	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem		
							F1901	Additional Surgery	A24	No Apparent Device Problem		
					CT3002	Implantable synthetic polymer	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem		
							F1901	Additional Surgery	A24	No Apparent Device Problem		
					CT3006	Implantable inorganic material	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem		
							F1901	Additional Surgery	A24	No Apparent Device Problem		
					CT314	Electrically powered	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem		
							F1901	Additional Surgery	A24	No Apparent Device Problem		

					CT321	Long-term surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem		
							F1901	Additional Surgery	A24	No Apparent Device Problem		
					CT449	Generators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem		
							F1901	Additional Surgery	A24	No Apparent Device Problem		
					CT981	Single use	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem		
							F1901	Additional Surgery	A24	No Apparent Device Problem		
					CT983	Surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem		
							F1901	Additional Surgery	A24	No Apparent Device Problem		
11911	2022		47265	Dual-chamber implantable pacemaker, rate-responsive	CT141	Electrodes/Leads and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	02/12/2022	02/12/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	02/12/2022	02/12/2022
					CT145	Cardiovascular devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	02/12/2022	02/12/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	02/12/2022	02/12/2022
					CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	02/12/2022	02/12/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	02/12/2022	02/12/2022
					CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	02/12/2022	02/12/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	02/12/2022	02/12/2022
					CT233	Surgical	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	02/12/2022	02/12/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	02/12/2022	02/12/2022
					CT2406	Implantable	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	02/12/2022	02/12/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	02/12/2022	02/12/2022
					CT2407	Active implantable	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	02/12/2022	02/12/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	02/12/2022	02/12/2022
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	02/12/2022	02/12/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	02/12/2022	02/12/2022
					CT2727	Active implant devices/ancillaries	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	02/12/2022	02/12/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	02/12/2022	02/12/2022
					CT2750	Implantable cardiac electrical stimulators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	02/12/2022	02/12/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	02/12/2022	02/12/2022
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	02/12/2022	02/12/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	02/12/2022	02/12/2022
					CT2985	Active	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	02/12/2022	02/12/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	02/12/2022	02/12/2022
					CT3001	Implantable metal	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	02/12/2022	02/12/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	02/12/2022	02/12/2022
					CT3002	Implantable synthetic polymer	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	02/12/2022	02/12/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	02/12/2022	02/12/2022
					CT3006	Implantable inorganic material	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	02/12/2022	02/12/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	02/12/2022	02/12/2022
					CT314	Electrically powered	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	02/12/2022	02/12/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	02/12/2022	02/12/2022
					CT321	Long-term surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	02/12/2022	02/12/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	02/12/2022	02/12/2022
					CT449	Generators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	02/12/2022	02/12/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	02/12/2022	02/12/2022
					CT981	Single use	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	02/12/2022	02/12/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	02/12/2022	02/12/2022
					CT983	Surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	02/12/2022	02/12/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	02/12/2022	02/12/2022
11912	2022	All other reportable incidents	47265	Dual-chamber implantable pacemaker, rate-responsive	CT141	Electrodes/Leads and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	20/12/2022	20/12/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	20/12/2022	20/12/2022
					CT145	Cardiovascular devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	20/12/2022	20/12/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	20/12/2022	20/12/2022
					CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	20/12/2022	20/12/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	20/12/2022	20/12/2022
					CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	20/12/2022	20/12/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	20/12/2022	20/12/2022
					CT233	Surgical	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	20/12/2022	20/12/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	20/12/2022	20/12/2022
					CT2406	Implantable	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	20/12/2022	20/12/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	20/12/2022	20/12/2022
					CT2407	Active implantable	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	20/12/2022	20/12/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	20/12/2022	20/12/2022
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	20/12/2022	20/12/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	20/12/2022	20/12/2022
					CT2727	Active implant devices/ancillaries	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	20/12/2022	20/12/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	20/12/2022	20/12/2022
					CT2750	Implantable cardiac electrical stimulators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	20/12/2022	20/12/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	20/12/2022	20/12/2022
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	20/12/2022	20/12/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	20/12/2022	20/12/2022
					CT2985	Active	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	20/12/2022	20/12/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	20/12/2022	20/12/2022
					CT3001	Implantable metal	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	20/12/2022	20/12/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	20/12/2022	20/12/2022
					CT3002	Implantable synthetic polymer	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	20/12/2022	20/12/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	20/12/2022	20/12/2022
					CT3006	Implantable inorganic material	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	20/12/2022	20/12/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	20/12/2022	20/12/2022
					CT314	Electrically powered	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	20/12/2022	20/12/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	20/12/2022	20/12/2022
					CT321	Long-term surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	20/12/2022	20/12/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	20/12/2022	20/12/2022
					CT449	Generators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	20/12/2022	20/12/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	20/12/2022	20/12/2022
					CT981	Single use	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	20/12/2022	20/12/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	20/12/2022	20/12/2022
					CT983	Surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	20/12/2022	20/12/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	20/12/2022	20/12/2022

11913	2020	Unanticipated serious deterioration in state of health	35224	External pacemaker, epicardial pacing	CT1684	Cardiac pacemakers and associated devices	E0602	Cardiac Arrest	A0712	Pacing Problem		
							F2306	Resuscitation	A0712	Pacing Problem		
					CT1685	Cardiac pacemakers	E0602	Cardiac Arrest	A0712	Pacing Problem		
							F2306	Resuscitation	A0712	Pacing Problem		
11914	2020	All other reportable incidents	35224	External pacemaker, epicardial pacing	CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A210104	Unclear Information		
									A2105	Inadequate or Insufficient Training		
					CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A210104	Unclear Information		
									A2105	Inadequate or Insufficient Training		
11915	2021	Unanticipated serious deterioration in state of health	35224	External pacemaker, epicardial pacing	CT1684	Cardiac pacemakers and associated devices						
					CT1685	Cardiac pacemakers						
11916	2022	All other reportable incidents	46717	Pacing/defibrillation lead stylet	CT1684	Cardiac pacemakers and associated devices	E200801	Device Embedded In Tissue or Plaque	A0401	Break		
							E2403	No Clinical Signs, Symptoms or Conditions	A0401	Break		
							F15	Recognized Device or Procedural Complication	A0401	Break		
11917	2019	All other reportable incidents	44863	Temporary cardiac pacing/stimulation connection cable, single-use	CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A0501	Detachment of Device or Device Component		
11918	2021	All other reportable incidents	35224	External pacemaker, epicardial pacing	CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions				
							F26	No Health Consequences or Impact				
					CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions				
							F26	No Health Consequences or Impact				
11919	2022	All other reportable incidents	17882	Physiologic-monitoring defibrillation system	CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A071301	Failure to Deliver Shock/Stimulation		
							F27	Problem Identified During Non-Clinical Procedure	A071301	Failure to Deliver Shock/Stimulation		
11920	2022	All other reportable incidents	17882	Physiologic-monitoring defibrillation system	CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A090811	Unable to Obtain Readings		
							F27	Problem Identified During Non-Clinical Procedure	A090811	Unable to Obtain Readings		
11921	2022	All other reportable incidents	35852	Single-chamber implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
11922	2022	All other reportable incidents	17882	Physiologic-monitoring defibrillation system	CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070803	Failure to Power Up		
							F27	Problem Identified During Non-Clinical Procedure	A070803	Failure to Power Up		
11923	2022	All other reportable incidents	47265	Dual-chamber implantable pacemaker, rate-responsive	CT141	Electrodes/Leads and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	29/11/2022	29/11/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	29/11/2022	29/11/2022
					CT145	Cardiovascular devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	29/11/2022	29/11/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	29/11/2022	29/11/2022
					CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	29/11/2022	29/11/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	29/11/2022	29/11/2022
					CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	29/11/2022	29/11/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	29/11/2022	29/11/2022
					CT233	Surgical	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	29/11/2022	29/11/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	29/11/2022	29/11/2022
					CT2406	Implantable	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	29/11/2022	29/11/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	29/11/2022	29/11/2022
					CT2407	Active implantable	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	29/11/2022	29/11/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	29/11/2022	29/11/2022
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	29/11/2022	29/11/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	29/11/2022	29/11/2022
					CT2727	Active implant devices/ancillaries	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	29/11/2022	29/11/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	29/11/2022	29/11/2022
					CT2750	Implantable cardiac electrical stimulators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	29/11/2022	29/11/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	29/11/2022	29/11/2022
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	29/11/2022	29/11/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	29/11/2022	29/11/2022
					CT2985	Active	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	29/11/2022	29/11/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	29/11/2022	29/11/2022
					CT3001	Implantable metal	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	29/11/2022	29/11/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	29/11/2022	29/11/2022
					CT3002	Implantable synthetic polymer	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	29/11/2022	29/11/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	29/11/2022	29/11/2022
					CT3006	Implantable inorganic material	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	29/11/2022	29/11/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	29/11/2022	29/11/2022
					CT314	Electrically powered	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	29/11/2022	29/11/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	29/11/2022	29/11/2022
					CT321	Long-term surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	29/11/2022	29/11/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	29/11/2022	29/11/2022
					CT449	Generators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	29/11/2022	29/11/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	29/11/2022	29/11/2022
					CT981	Single use	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	29/11/2022	29/11/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	29/11/2022	29/11/2022
					CT983	Surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	29/11/2022	29/11/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	29/11/2022	29/11/2022
11924	2022	All other reportable incidents	47265	Dual-chamber implantable pacemaker, rate-responsive	CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	05/12/2022	05/12/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	05/12/2022	05/12/2022
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	05/12/2022	05/12/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	05/12/2022	05/12/2022
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	05/12/2022	05/12/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	05/12/2022	05/12/2022
11925	2022	All other reportable incidents	47265	Dual-chamber implantable pacemaker, rate-responsive	CT141	Electrodes/Leads and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	20/09/2022	20/09/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	20/09/2022	20/09/2022
					CT145	Cardiovascular devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	20/09/2022	20/09/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	20/09/2022	20/09/2022
					CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	20/09/2022	20/09/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	20/09/2022	20/09/2022
					CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	20/09/2022	20/09/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	20/09/2022	20/09/2022
					CT233	Surgical	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	20/09/2022	20/09/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	20/09/2022	20/09/2022
					CT2406	Implantable	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	20/09/2022	20/09/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	20/09/2022	20/09/2022
					CT2407	Active implantable	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	20/09/2022	20/09/2022

							F1901	Additional Surgery	A24	No Apparent Device Problem	20/09/2022	20/09/2022
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	20/09/2022	20/09/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	20/09/2022	20/09/2022
					CT2727	Active implant devices/ancillaries	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	20/09/2022	20/09/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	20/09/2022	20/09/2022
					CT2750	Implantable cardiac electrical stimulators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	20/09/2022	20/09/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	20/09/2022	20/09/2022
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	20/09/2022	20/09/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	20/09/2022	20/09/2022
					CT2985	Active	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	20/09/2022	20/09/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	20/09/2022	20/09/2022
					CT3001	Implantable metal	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	20/09/2022	20/09/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	20/09/2022	20/09/2022
					CT3002	Implantable synthetic polymer	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	20/09/2022	20/09/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	20/09/2022	20/09/2022
					CT3006	Implantable inorganic material	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	20/09/2022	20/09/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	20/09/2022	20/09/2022
					CT314	Electrically powered	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	20/09/2022	20/09/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	20/09/2022	20/09/2022
					CT321	Long-term surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	20/09/2022	20/09/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	20/09/2022	20/09/2022
					CT449	Generators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	20/09/2022	20/09/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	20/09/2022	20/09/2022
					CT981	Single use	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	20/09/2022	20/09/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	20/09/2022	20/09/2022
					CT983	Surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	20/09/2022	20/09/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	20/09/2022	20/09/2022
11926	2022		47265	Dual-chamber implantable pacemaker, rate-responsive	CT141	Electrodes/Leads and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem		
							F1901	Additional Surgery	A24	No Apparent Device Problem		
					CT145	Cardiovascular devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem		
							F1901	Additional Surgery	A24	No Apparent Device Problem		
					CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem		
							F1901	Additional Surgery	A24	No Apparent Device Problem		
					CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem		
							F1901	Additional Surgery	A24	No Apparent Device Problem		
					CT233	Surgical	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem		
							F1901	Additional Surgery	A24	No Apparent Device Problem		
					CT2406	Implantable	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem		
							F1901	Additional Surgery	A24	No Apparent Device Problem		
					CT2407	Active implantable	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem		
							F1901	Additional Surgery	A24	No Apparent Device Problem		
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem		
							F1901	Additional Surgery	A24	No Apparent Device Problem		
					CT2727	Active implant devices/ancillaries	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem		
							F1901	Additional Surgery	A24	No Apparent Device Problem		
					CT2750	Implantable cardiac electrical stimulators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem		
							F1901	Additional Surgery	A24	No Apparent Device Problem		
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem		
							F1901	Additional Surgery	A24	No Apparent Device Problem		
					CT2985	Active	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem		
							F1901	Additional Surgery	A24	No Apparent Device Problem		
					CT3001	Implantable metal	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem		
							F1901	Additional Surgery	A24	No Apparent Device Problem		
					CT3002	Implantable synthetic polymer	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem		
							F1901	Additional Surgery	A24	No Apparent Device Problem		
					CT3006	Implantable inorganic material	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem		
							F1901	Additional Surgery	A24	No Apparent Device Problem		
					CT314	Electrically powered	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem		
							F1901	Additional Surgery	A24	No Apparent Device Problem		
					CT321	Long-term surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem		
							F1901	Additional Surgery	A24	No Apparent Device Problem		
					CT449	Generators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem		
							F1901	Additional Surgery	A24	No Apparent Device Problem		
					CT981	Single use	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem		
							F1901	Additional Surgery	A24	No Apparent Device Problem		
					CT983	Surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem		
							F1901	Additional Surgery	A24	No Apparent Device Problem		
11927	2022	All other reportable incidents	17882	Physiologic-monitoring defibrillation system	CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A0719	Unexpected Shutdown		
							F27	Problem Identified During Non-Clinical Procedure	A0719	Unexpected Shutdown		
11928	2022	All other reportable incidents	35852	Single-chamber implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E2104	Electric Shock	A071302	Inappropriate/Inadequate Shock/Stimulation		
							F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A071302	Inappropriate/Inadequate Shock/Stimulation		
11929	2022	All other reportable incidents	47265	Dual-chamber implantable pacemaker, rate-responsive	CT141	Electrodes/Leads and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A1606	Premature Indicator Activation		
							F1905	Device Revision or Replacement	A1606	Premature Indicator Activation		
					CT145	Cardiovascular devices	E2403	No Clinical Signs, Symptoms or Conditions	A1606	Premature Indicator Activation		
							F1905	Device Revision or Replacement	A1606	Premature Indicator Activation		
					CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A1606	Premature Indicator Activation		
							F1905	Device Revision or Replacement	A1606	Premature Indicator Activation		
					CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A1606	Premature Indicator Activation		
							F1905	Device Revision or Replacement	A1606	Premature Indicator Activation		
					CT233	Surgical	E2403	No Clinical Signs, Symptoms or Conditions	A1606	Premature Indicator Activation		
							F1905	Device Revision or Replacement	A1606	Premature Indicator Activation		
					CT2406	Implantable	E2403	No Clinical Signs, Symptoms or Conditions	A1606	Premature Indicator Activation		
							F1905	Device Revision or Replacement	A1606	Premature Indicator Activation		
					CT2407	Active implantable	E2403	No Clinical Signs, Symptoms or Conditions	A1606	Premature Indicator Activation		
							F1905	Device Revision or Replacement	A1606	Premature Indicator Activation		
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A1606	Premature Indicator Activation		
							F1905	Device Revision or Replacement	A1606	Premature Indicator Activation		
					CT2727	Active implant devices/ancillaries	E2403	No Clinical Signs, Symptoms or Conditions	A1606	Premature Indicator Activation		

							F1905	Device Revision or Replacement	A1606	Premature Indicator Activation		
					CT2750	Implantable cardiac electrical stimulators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A1606	Premature Indicator Activation		
							F1905	Device Revision or Replacement	A1606	Premature Indicator Activation		
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A1606	Premature Indicator Activation		
							F1905	Device Revision or Replacement	A1606	Premature Indicator Activation		
					CT2985	Active	E2403	No Clinical Signs, Symptoms or Conditions	A1606	Premature Indicator Activation		
							F1905	Device Revision or Replacement	A1606	Premature Indicator Activation		
					CT3001	Implantable metal	E2403	No Clinical Signs, Symptoms or Conditions	A1606	Premature Indicator Activation		
							F1905	Device Revision or Replacement	A1606	Premature Indicator Activation		
					CT3002	Implantable synthetic polymer	E2403	No Clinical Signs, Symptoms or Conditions	A1606	Premature Indicator Activation		
							F1905	Device Revision or Replacement	A1606	Premature Indicator Activation		
					CT3006	Implantable inorganic material	E2403	No Clinical Signs, Symptoms or Conditions	A1606	Premature Indicator Activation		
							F1905	Device Revision or Replacement	A1606	Premature Indicator Activation		
					CT314	Electrically powered	E2403	No Clinical Signs, Symptoms or Conditions	A1606	Premature Indicator Activation		
							F1905	Device Revision or Replacement	A1606	Premature Indicator Activation		
					CT321	Long-term surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A1606	Premature Indicator Activation		
							F1905	Device Revision or Replacement	A1606	Premature Indicator Activation		
					CT449	Generators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A1606	Premature Indicator Activation		
							F1905	Device Revision or Replacement	A1606	Premature Indicator Activation		
					CT981	Single use	E2403	No Clinical Signs, Symptoms or Conditions	A1606	Premature Indicator Activation		
							F1905	Device Revision or Replacement	A1606	Premature Indicator Activation		
					CT983	Surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A1606	Premature Indicator Activation		
							F1905	Device Revision or Replacement	A1606	Premature Indicator Activation		
11930	2022		47265	Dual-chamber implantable pacemaker, rate-responsive	CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem		
									A09	Output Problem		
							F26	No Health Consequences or Impact	A0705	Battery Problem		
									A09	Output Problem		
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem		
									A09	Output Problem		
							F26	No Health Consequences or Impact	A0705	Battery Problem		
									A09	Output Problem		
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem		
									A09	Output Problem		
							F26	No Health Consequences or Impact	A0705	Battery Problem		
									A09	Output Problem		
11931	2022				CT257	Implantable cardiac pacemakers	F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A07	Electrical /Electronic Property Problem		
									A0712	Pacing Problem		
11932	2022	All other reportable incidents	47267	Single-chamber implantable pacemaker, rate-responsive	CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	02/11/2022	02/11/2022
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	02/11/2022	02/11/2022
							F1903	Device Explantation	A070504	Premature Discharge of Battery	02/11/2022	02/11/2022
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	02/11/2022	02/11/2022
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	02/11/2022	02/11/2022
							F1903	Device Explantation	A070504	Premature Discharge of Battery	02/11/2022	02/11/2022
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	02/11/2022	02/11/2022
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	02/11/2022	02/11/2022
							F1903	Device Explantation	A070504	Premature Discharge of Battery	02/11/2022	02/11/2022
11933	2023		47265	Dual-chamber implantable pacemaker, rate-responsive	CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem		
							F1901	Additional Surgery	A24	No Apparent Device Problem		
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem		
							F1901	Additional Surgery	A24	No Apparent Device Problem		
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem		
							F1901	Additional Surgery	A24	No Apparent Device Problem		
11934	2023		47265	Dual-chamber implantable pacemaker, rate-responsive	CT141	Electrodes/Leads and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem		
							F1901	Additional Surgery	A24	No Apparent Device Problem		
					CT145	Cardiovascular devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem		
							F1901	Additional Surgery	A24	No Apparent Device Problem		
					CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem		
							F1901	Additional Surgery	A24	No Apparent Device Problem		
					CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem		
							F1901	Additional Surgery	A24	No Apparent Device Problem		
					CT233	Surgical	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem		
							F1901	Additional Surgery	A24	No Apparent Device Problem		
					CT2406	Implantable	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem		
							F1901	Additional Surgery	A24	No Apparent Device Problem		
					CT2407	Active implantable	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem		
							F1901	Additional Surgery	A24	No Apparent Device Problem		
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem		
							F1901	Additional Surgery	A24	No Apparent Device Problem		
					CT2727	Active implant devices/ancillaries	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem		
							F1901	Additional Surgery	A24	No Apparent Device Problem		
					CT2750	Implantable cardiac electrical stimulators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem		
							F1901	Additional Surgery	A24	No Apparent Device Problem		
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem		
							F1901	Additional Surgery	A24	No Apparent Device Problem		
					CT2985	Active	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem		
							F1901	Additional Surgery	A24	No Apparent Device Problem		
					CT3001	Implantable metal	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem		
							F1901	Additional Surgery	A24	No Apparent Device Problem		
					CT3002	Implantable synthetic polymer	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem		
							F1901	Additional Surgery	A24	No Apparent Device Problem		
					CT3006	Implantable inorganic material	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem		
							F1901	Additional Surgery	A24	No Apparent Device Problem		
					CT314	Electrically powered	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem		
							F1901	Additional Surgery	A24	No Apparent Device Problem		
					CT321	Long-term surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem		
							F1901	Additional Surgery	A24	No Apparent Device Problem		
					CT449	Generators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem		
							F1901	Additional Surgery	A24	No Apparent Device Problem		

					CT981	Single use	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem		
							F1901	Additional Surgery	A24	No Apparent Device Problem		
					CT983	Surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem		
							F1901	Additional Surgery	A24	No Apparent Device Problem		
11935	2023		47265	Dual-chamber implantable pacemaker, rate-responsive	CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem		
							F1901	Additional Surgery	A24	No Apparent Device Problem		
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem		
							F1901	Additional Surgery	A24	No Apparent Device Problem		
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem		
							F1901	Additional Surgery	A24	No Apparent Device Problem		
11936	2023		47265	Dual-chamber implantable pacemaker, rate-responsive	CT141	Electrodes/Leads and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	05/12/2022	05/12/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	05/12/2022	05/12/2022
					CT145	Cardiovascular devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	05/12/2022	05/12/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	05/12/2022	05/12/2022
					CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	05/12/2022	05/12/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	05/12/2022	05/12/2022
					CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	05/12/2022	05/12/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	05/12/2022	05/12/2022
					CT233	Surgical	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	05/12/2022	05/12/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	05/12/2022	05/12/2022
					CT2406	Implantable	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	05/12/2022	05/12/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	05/12/2022	05/12/2022
					CT2407	Active implantable	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	05/12/2022	05/12/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	05/12/2022	05/12/2022
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	05/12/2022	05/12/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	05/12/2022	05/12/2022
					CT2727	Active implant devices/ancillaries	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	05/12/2022	05/12/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	05/12/2022	05/12/2022
					CT2750	Implantable cardiac electrical stimulators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	05/12/2022	05/12/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	05/12/2022	05/12/2022
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	05/12/2022	05/12/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	05/12/2022	05/12/2022
					CT2985	Active	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	05/12/2022	05/12/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	05/12/2022	05/12/2022
					CT3001	Implantable metal	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	05/12/2022	05/12/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	05/12/2022	05/12/2022
					CT3002	Implantable synthetic polymer	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	05/12/2022	05/12/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	05/12/2022	05/12/2022
					CT3006	Implantable inorganic material	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	05/12/2022	05/12/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	05/12/2022	05/12/2022
					CT314	Electrically powered	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	05/12/2022	05/12/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	05/12/2022	05/12/2022
					CT321	Long-term surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	05/12/2022	05/12/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	05/12/2022	05/12/2022
					CT449	Generators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	05/12/2022	05/12/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	05/12/2022	05/12/2022
					CT981	Single use	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	05/12/2022	05/12/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	05/12/2022	05/12/2022
					CT983	Surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	05/12/2022	05/12/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	05/12/2022	05/12/2022
11937	2023	All other reportable incidents	35223	Endocardial/interventricular septal pacing lead	CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070909	Oversensing		
							F1901	Additional Surgery	A070909	Oversensing		
11938	2023		17882	Physiologic-monitoring defibrillation system	CT1684	Cardiac pacemakers and associated devices	F26	No Health Consequences or Impact	A110201	Application Program Freezes or Fails to Launch		
11939	2022	All other reportable incidents	35852	Single-chamber implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	03/02/2023	03/02/2023
									A0908	Incorrect, Inadequate or Imprecise Result or Readings	03/02/2023	03/02/2023
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	03/02/2023	03/02/2023
									A0908	Incorrect, Inadequate or Imprecise Result or Readings	03/02/2023	03/02/2023
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	03/02/2023	03/02/2023
									A0908	Incorrect, Inadequate or Imprecise Result or Readings	03/02/2023	03/02/2023
11940	2023	All other reportable incidents	37265	Dual-chamber implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A070909	Oversensing		
									A070910	Undersensing		
									A0712	Pacing Problem		
							F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A070909	Oversensing		
									A070910	Undersensing		
									A0712	Pacing Problem		
11941	2023		47265	Dual-chamber implantable pacemaker, rate-responsive	CT141	Electrodes/Leads and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	28/11/2022	28/11/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	28/11/2022	28/11/2022
					CT145	Cardiovascular devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	28/11/2022	28/11/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	28/11/2022	28/11/2022
					CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	28/11/2022	28/11/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	28/11/2022	28/11/2022
					CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	28/11/2022	28/11/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	28/11/2022	28/11/2022
					CT233	Surgical	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	28/11/2022	28/11/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	28/11/2022	28/11/2022
					CT2406	Implantable	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	28/11/2022	28/11/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	28/11/2022	28/11/2022
					CT2407	Active implantable	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	28/11/2022	28/11/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	28/11/2022	28/11/2022
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	28/11/2022	28/11/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	28/11/2022	28/11/2022
					CT2727	Active implant devices/ancillaries	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	28/11/2022	28/11/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	28/11/2022	28/11/2022
					CT2750	Implantable cardiac electrical stimulators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	28/11/2022	28/11/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	28/11/2022	28/11/2022

					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	28/11/2022	28/11/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	28/11/2022	28/11/2022
					CT2985	Active	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	28/11/2022	28/11/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	28/11/2022	28/11/2022
					CT3001	Implantable metal	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	28/11/2022	28/11/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	28/11/2022	28/11/2022
					CT3002	Implantable synthetic polymer	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	28/11/2022	28/11/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	28/11/2022	28/11/2022
					CT3006	Implantable inorganic material	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	28/11/2022	28/11/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	28/11/2022	28/11/2022
					CT314	Electrically powered	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	28/11/2022	28/11/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	28/11/2022	28/11/2022
					CT321	Long-term surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	28/11/2022	28/11/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	28/11/2022	28/11/2022
					CT449	Generators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	28/11/2022	28/11/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	28/11/2022	28/11/2022
					CT981	Single use	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	28/11/2022	28/11/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	28/11/2022	28/11/2022
					CT983	Surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	28/11/2022	28/11/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	28/11/2022	28/11/2022
11942	2023		47265	Dual-chamber implantable pacemaker, rate-responsive	CT141	Electrodes/Leads and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem		
							F1901	Additional Surgery	A24	No Apparent Device Problem		
					CT145	Cardiovascular devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem		
							F1901	Additional Surgery	A24	No Apparent Device Problem		
					CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem		
							F1901	Additional Surgery	A24	No Apparent Device Problem		
					CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem		
							F1901	Additional Surgery	A24	No Apparent Device Problem		
					CT233	Surgical	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem		
							F1901	Additional Surgery	A24	No Apparent Device Problem		
					CT2406	Implantable	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem		
							F1901	Additional Surgery	A24	No Apparent Device Problem		
					CT2407	Active implantable	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem		
							F1901	Additional Surgery	A24	No Apparent Device Problem		
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem		
							F1901	Additional Surgery	A24	No Apparent Device Problem		
					CT2727	Active implant devices/ancillaries	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem		
							F1901	Additional Surgery	A24	No Apparent Device Problem		
					CT2750	Implantable cardiac electrical stimulators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem		
							F1901	Additional Surgery	A24	No Apparent Device Problem		
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem		
							F1901	Additional Surgery	A24	No Apparent Device Problem		
					CT2985	Active	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem		
							F1901	Additional Surgery	A24	No Apparent Device Problem		
					CT3001	Implantable metal	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem		
							F1901	Additional Surgery	A24	No Apparent Device Problem		
					CT3002	Implantable synthetic polymer	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem		
							F1901	Additional Surgery	A24	No Apparent Device Problem		
					CT3006	Implantable inorganic material	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem		
							F1901	Additional Surgery	A24	No Apparent Device Problem		
					CT314	Electrically powered	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem		
							F1901	Additional Surgery	A24	No Apparent Device Problem		
					CT321	Long-term surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem		
							F1901	Additional Surgery	A24	No Apparent Device Problem		
					CT449	Generators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem		
							F1901	Additional Surgery	A24	No Apparent Device Problem		
					CT981	Single use	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem		
							F1901	Additional Surgery	A24	No Apparent Device Problem		
					CT983	Surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem		
							F1901	Additional Surgery	A24	No Apparent Device Problem		
11943	2023		47265	Dual-chamber implantable pacemaker, rate-responsive	CT141	Electrodes/Leads and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	14/10/2022	14/10/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	14/10/2022	14/10/2022
					CT145	Cardiovascular devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	14/10/2022	14/10/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	14/10/2022	14/10/2022
					CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	14/10/2022	14/10/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	14/10/2022	14/10/2022
					CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	14/10/2022	14/10/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	14/10/2022	14/10/2022
					CT233	Surgical	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	14/10/2022	14/10/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	14/10/2022	14/10/2022
					CT2406	Implantable	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	14/10/2022	14/10/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	14/10/2022	14/10/2022
					CT2407	Active implantable	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	14/10/2022	14/10/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	14/10/2022	14/10/2022
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	14/10/2022	14/10/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	14/10/2022	14/10/2022
					CT2727	Active implant devices/ancillaries	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	14/10/2022	14/10/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	14/10/2022	14/10/2022
					CT2750	Implantable cardiac electrical stimulators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	14/10/2022	14/10/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	14/10/2022	14/10/2022
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	14/10/2022	14/10/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	14/10/2022	14/10/2022
					CT2985	Active	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	14/10/2022	14/10/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	14/10/2022	14/10/2022
					CT3001	Implantable metal	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	14/10/2022	14/10/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	14/10/2022	14/10/2022
					CT3002	Implantable synthetic polymer	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	14/10/2022	14/10/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	14/10/2022	14/10/2022

					CT3006	Implantable inorganic material	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	14/10/2022	14/10/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	14/10/2022	14/10/2022
					CT314	Electrically powered	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	14/10/2022	14/10/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	14/10/2022	14/10/2022
					CT321	Long-term surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	14/10/2022	14/10/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	14/10/2022	14/10/2022
					CT449	Generators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	14/10/2022	14/10/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	14/10/2022	14/10/2022
					CT981	Single use	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	14/10/2022	14/10/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	14/10/2022	14/10/2022
					CT983	Surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	14/10/2022	14/10/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	14/10/2022	14/10/2022
11944	2023		47265	Dual-chamber implantable pacemaker, rate-responsive	CT141	Electrodes/Leads and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	18/11/2022	18/11/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	18/11/2022	18/11/2022
					CT145	Cardiovascular devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	18/11/2022	18/11/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	18/11/2022	18/11/2022
					CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	18/11/2022	18/11/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	18/11/2022	18/11/2022
					CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	18/11/2022	18/11/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	18/11/2022	18/11/2022
					CT233	Surgical	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	18/11/2022	18/11/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	18/11/2022	18/11/2022
					CT2406	Implantable	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	18/11/2022	18/11/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	18/11/2022	18/11/2022
					CT2407	Active implantable	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	18/11/2022	18/11/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	18/11/2022	18/11/2022
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	18/11/2022	18/11/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	18/11/2022	18/11/2022
					CT2727	Active implant devices/ancillaries	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	18/11/2022	18/11/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	18/11/2022	18/11/2022
					CT2750	Implantable cardiac electrical stimulators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	18/11/2022	18/11/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	18/11/2022	18/11/2022
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	18/11/2022	18/11/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	18/11/2022	18/11/2022
					CT2985	Active	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	18/11/2022	18/11/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	18/11/2022	18/11/2022
					CT3001	Implantable metal	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	18/11/2022	18/11/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	18/11/2022	18/11/2022
					CT3002	Implantable synthetic polymer	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	18/11/2022	18/11/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	18/11/2022	18/11/2022
					CT3006	Implantable inorganic material	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	18/11/2022	18/11/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	18/11/2022	18/11/2022
					CT314	Electrically powered	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	18/11/2022	18/11/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	18/11/2022	18/11/2022
					CT321	Long-term surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	18/11/2022	18/11/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	18/11/2022	18/11/2022
					CT449	Generators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	18/11/2022	18/11/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	18/11/2022	18/11/2022
					CT981	Single use	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	18/11/2022	18/11/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	18/11/2022	18/11/2022
					CT983	Surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	18/11/2022	18/11/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	18/11/2022	18/11/2022
11945	2023		47265	Dual-chamber implantable pacemaker, rate-responsive	CT141	Electrodes/Leads and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	09/11/2022	09/11/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	09/11/2022	09/11/2022
					CT145	Cardiovascular devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	09/11/2022	09/11/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	09/11/2022	09/11/2022
					CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	09/11/2022	09/11/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	09/11/2022	09/11/2022
					CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	09/11/2022	09/11/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	09/11/2022	09/11/2022
					CT233	Surgical	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	09/11/2022	09/11/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	09/11/2022	09/11/2022
					CT2406	Implantable	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	09/11/2022	09/11/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	09/11/2022	09/11/2022
					CT2407	Active implantable	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	09/11/2022	09/11/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	09/11/2022	09/11/2022
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	09/11/2022	09/11/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	09/11/2022	09/11/2022
					CT2727	Active implant devices/ancillaries	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	09/11/2022	09/11/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	09/11/2022	09/11/2022
					CT2750	Implantable cardiac electrical stimulators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	09/11/2022	09/11/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	09/11/2022	09/11/2022
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	09/11/2022	09/11/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	09/11/2022	09/11/2022
					CT2985	Active	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	09/11/2022	09/11/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	09/11/2022	09/11/2022
					CT3001	Implantable metal	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	09/11/2022	09/11/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	09/11/2022	09/11/2022
					CT3002	Implantable synthetic polymer	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	09/11/2022	09/11/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	09/11/2022	09/11/2022
					CT3006	Implantable inorganic material	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	09/11/2022	09/11/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	09/11/2022	09/11/2022
					CT314	Electrically powered	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	09/11/2022	09/11/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	09/11/2022	09/11/2022
					CT321	Long-term surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	09/11/2022	09/11/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	09/11/2022	09/11/2022
					CT449	Generators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	09/11/2022	09/11/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	09/11/2022	09/11/2022

					CT981	Single use	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	09/11/2022	09/11/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	09/11/2022	09/11/2022
					CT983	Surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	09/11/2022	09/11/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	09/11/2022	09/11/2022
11946	2023	All other reportable incidents	35223	Endocardial/interventricular septal pacing lead	CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A0401	Break	22/12/2022	22/12/2022
									A0701	Capturing Problem	22/12/2022	22/12/2022
									A0709	Device Sensing Problem	22/12/2022	22/12/2022
							F1901	Additional Surgery	A0401	Break	22/12/2022	22/12/2022
									A0701	Capturing Problem	22/12/2022	22/12/2022
									A0709	Device Sensing Problem	22/12/2022	22/12/2022
11947	2023	All other reportable incidents	35223	Endocardial/interventricular septal pacing lead	CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A0401	Break	22/12/2022	22/12/2022
									A070101	Failure to Capture	22/12/2022	22/12/2022
									A072202	Low Impedance	22/12/2022	22/12/2022
							F1901	Additional Surgery	A0401	Break	22/12/2022	22/12/2022
									A070101	Failure to Capture	22/12/2022	22/12/2022
									A072202	Low Impedance	22/12/2022	22/12/2022
11948	2023		47265	Dual-chamber implantable pacemaker, rate-responsive	CT141	Electrodes/Leads and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem		
							F26	No Health Consequences or Impact	A24	No Apparent Device Problem		
					CT145	Cardiovascular devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem		
							F26	No Health Consequences or Impact	A24	No Apparent Device Problem		
					CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem		
							F26	No Health Consequences or Impact	A24	No Apparent Device Problem		
					CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem		
							F26	No Health Consequences or Impact	A24	No Apparent Device Problem		
					CT233	Surgical	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem		
							F26	No Health Consequences or Impact	A24	No Apparent Device Problem		
					CT2406	Implantable	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem		
							F26	No Health Consequences or Impact	A24	No Apparent Device Problem		
					CT2407	Active implantable	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem		
							F26	No Health Consequences or Impact	A24	No Apparent Device Problem		
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem		
							F26	No Health Consequences or Impact	A24	No Apparent Device Problem		
					CT2727	Active implant devices/ancillaries	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem		
							F26	No Health Consequences or Impact	A24	No Apparent Device Problem		
					CT2750	Implantable cardiac electrical stimulators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem		
							F26	No Health Consequences or Impact	A24	No Apparent Device Problem		
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem		
							F26	No Health Consequences or Impact	A24	No Apparent Device Problem		
					CT2985	Active	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem		
							F26	No Health Consequences or Impact	A24	No Apparent Device Problem		
					CT3001	Implantable metal	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem		
							F26	No Health Consequences or Impact	A24	No Apparent Device Problem		
					CT3002	Implantable synthetic polymer	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem		
							F26	No Health Consequences or Impact	A24	No Apparent Device Problem		
					CT3006	Implantable inorganic material	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem		
							F26	No Health Consequences or Impact	A24	No Apparent Device Problem		
					CT314	Electrically powered	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem		
							F26	No Health Consequences or Impact	A24	No Apparent Device Problem		
					CT321	Long-term surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem		
							F26	No Health Consequences or Impact	A24	No Apparent Device Problem		
					CT449	Generators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem		
							F26	No Health Consequences or Impact	A24	No Apparent Device Problem		
					CT981	Single use	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem		
							F26	No Health Consequences or Impact	A24	No Apparent Device Problem		
					CT983	Surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem		
							F26	No Health Consequences or Impact	A24	No Apparent Device Problem		
11949	2023		47265	Dual-chamber implantable pacemaker, rate-responsive	CT141	Electrodes/Leads and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	17/02/2023	17/02/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	17/02/2023	17/02/2023
					CT145	Cardiovascular devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	17/02/2023	17/02/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	17/02/2023	17/02/2023
					CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	17/02/2023	17/02/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	17/02/2023	17/02/2023
					CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	17/02/2023	17/02/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	17/02/2023	17/02/2023
					CT233	Surgical	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	17/02/2023	17/02/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	17/02/2023	17/02/2023
					CT2406	Implantable	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	17/02/2023	17/02/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	17/02/2023	17/02/2023
					CT2407	Active implantable	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	17/02/2023	17/02/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	17/02/2023	17/02/2023
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	17/02/2023	17/02/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	17/02/2023	17/02/2023
					CT2727	Active implant devices/ancillaries	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	17/02/2023	17/02/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	17/02/2023	17/02/2023
					CT2750	Implantable cardiac electrical stimulators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	17/02/2023	17/02/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	17/02/2023	17/02/2023
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	17/02/2023	17/02/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	17/02/2023	17/02/2023
					CT2985	Active	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	17/02/2023	17/02/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	17/02/2023	17/02/2023
					CT3001	Implantable metal	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	17/02/2023	17/02/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	17/02/2023	17/02/2023
					CT3002	Implantable synthetic polymer	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	17/02/2023	17/02/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	17/02/2023	17/02/2023
					CT3006	Implantable inorganic material	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	17/02/2023	17/02/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	17/02/2023	17/02/2023
					CT314	Electrically powered	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	17/02/2023	17/02/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	17/02/2023	17/02/2023

					CT321	Long-term surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	17/02/2023	17/02/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	17/02/2023	17/02/2023
					CT449	Generators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	17/02/2023	17/02/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	17/02/2023	17/02/2023
					CT981	Single use	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	17/02/2023	17/02/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	17/02/2023	17/02/2023
					CT983	Surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	17/02/2023	17/02/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	17/02/2023	17/02/2023
11950	2023	All other reportable incidents	35852	Single-chamber implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
11951	2023		47265	Dual-chamber implantable pacemaker, rate-responsive	CT141	Electrodes/Leads and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	30/11/2022	30/11/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	30/11/2022	30/11/2022
					CT145	Cardiovascular devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	30/11/2022	30/11/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	30/11/2022	30/11/2022
					CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	30/11/2022	30/11/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	30/11/2022	30/11/2022
					CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	30/11/2022	30/11/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	30/11/2022	30/11/2022
					CT233	Surgical	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	30/11/2022	30/11/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	30/11/2022	30/11/2022
					CT2406	Implantable	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	30/11/2022	30/11/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	30/11/2022	30/11/2022
					CT2407	Active implantable	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	30/11/2022	30/11/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	30/11/2022	30/11/2022
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	30/11/2022	30/11/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	30/11/2022	30/11/2022
					CT2727	Active implant devices/ancillaries	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	30/11/2022	30/11/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	30/11/2022	30/11/2022
					CT2750	Implantable cardiac electrical stimulators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	30/11/2022	30/11/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	30/11/2022	30/11/2022
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	30/11/2022	30/11/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	30/11/2022	30/11/2022
					CT2985	Active	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	30/11/2022	30/11/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	30/11/2022	30/11/2022
					CT3001	Implantable metal	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	30/11/2022	30/11/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	30/11/2022	30/11/2022
					CT3002	Implantable synthetic polymer	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	30/11/2022	30/11/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	30/11/2022	30/11/2022
					CT3006	Implantable inorganic material	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	30/11/2022	30/11/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	30/11/2022	30/11/2022
					CT314	Electrically powered	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	30/11/2022	30/11/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	30/11/2022	30/11/2022
					CT321	Long-term surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	30/11/2022	30/11/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	30/11/2022	30/11/2022
					CT449	Generators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	30/11/2022	30/11/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	30/11/2022	30/11/2022
					CT981	Single use	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	30/11/2022	30/11/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	30/11/2022	30/11/2022
					CT983	Surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	30/11/2022	30/11/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	30/11/2022	30/11/2022
11952	2023	All other reportable incidents	47265	Dual-chamber implantable pacemaker, rate-responsive	CT141	Electrodes/Leads and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	23/12/2022	23/12/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	23/12/2022	23/12/2022
					CT145	Cardiovascular devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	23/12/2022	23/12/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	23/12/2022	23/12/2022
					CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	23/12/2022	23/12/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	23/12/2022	23/12/2022
					CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	23/12/2022	23/12/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	23/12/2022	23/12/2022
					CT233	Surgical	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	23/12/2022	23/12/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	23/12/2022	23/12/2022
					CT2406	Implantable	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	23/12/2022	23/12/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	23/12/2022	23/12/2022
					CT2407	Active implantable	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	23/12/2022	23/12/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	23/12/2022	23/12/2022
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	23/12/2022	23/12/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	23/12/2022	23/12/2022
					CT2727	Active implant devices/ancillaries	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	23/12/2022	23/12/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	23/12/2022	23/12/2022
					CT2750	Implantable cardiac electrical stimulators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	23/12/2022	23/12/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	23/12/2022	23/12/2022
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	23/12/2022	23/12/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	23/12/2022	23/12/2022
					CT2985	Active	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	23/12/2022	23/12/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	23/12/2022	23/12/2022
					CT3001	Implantable metal	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	23/12/2022	23/12/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	23/12/2022	23/12/2022
					CT3002	Implantable synthetic polymer	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	23/12/2022	23/12/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	23/12/2022	23/12/2022
					CT3006	Implantable inorganic material	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	23/12/2022	23/12/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	23/12/2022	23/12/2022
					CT314	Electrically powered	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	23/12/2022	23/12/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	23/12/2022	23/12/2022
					CT321	Long-term surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	23/12/2022	23/12/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	23/12/2022	23/12/2022
					CT449	Generators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	23/12/2022	23/12/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	23/12/2022	23/12/2022
					CT981	Single use	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	23/12/2022	23/12/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	23/12/2022	23/12/2022

					CT983	Surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	23/12/2022	23/12/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	23/12/2022	23/12/2022
11953	2023		47265	Dual-chamber implantable pacemaker, rate-responsive	CT141	Electrodes/Leads and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	28/12/2022	28/12/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	28/12/2022	28/12/2022
					CT145	Cardiovascular devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	28/12/2022	28/12/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	28/12/2022	28/12/2022
					CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	28/12/2022	28/12/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	28/12/2022	28/12/2022
					CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	28/12/2022	28/12/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	28/12/2022	28/12/2022
					CT233	Surgical	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	28/12/2022	28/12/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	28/12/2022	28/12/2022
					CT2406	Implantable	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	28/12/2022	28/12/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	28/12/2022	28/12/2022
					CT2407	Active implantable	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	28/12/2022	28/12/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	28/12/2022	28/12/2022
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	28/12/2022	28/12/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	28/12/2022	28/12/2022
					CT2727	Active implant devices/ancillaries	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	28/12/2022	28/12/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	28/12/2022	28/12/2022
					CT2750	Implantable cardiac electrical stimulators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	28/12/2022	28/12/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	28/12/2022	28/12/2022
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	28/12/2022	28/12/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	28/12/2022	28/12/2022
					CT2985	Active	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	28/12/2022	28/12/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	28/12/2022	28/12/2022
					CT3001	Implantable metal	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	28/12/2022	28/12/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	28/12/2022	28/12/2022
					CT3002	Implantable synthetic polymer	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	28/12/2022	28/12/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	28/12/2022	28/12/2022
					CT3006	Implantable inorganic material	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	28/12/2022	28/12/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	28/12/2022	28/12/2022
					CT314	Electrically powered	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	28/12/2022	28/12/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	28/12/2022	28/12/2022
					CT321	Long-term surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	28/12/2022	28/12/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	28/12/2022	28/12/2022
					CT449	Generators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	28/12/2022	28/12/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	28/12/2022	28/12/2022
					CT981	Single use	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	28/12/2022	28/12/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	28/12/2022	28/12/2022
					CT983	Surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	28/12/2022	28/12/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	28/12/2022	28/12/2022
11954	2023		47270	Cardiac resynchronization therapy implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A25	No Apparent Adverse Event		
							F26	No Health Consequences or Impact	A25	No Apparent Adverse Event		
11955	2023	All other reportable incidents	35852	Single-chamber implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E060110	Ventricular Fibrillation	A0712	Pacing Problem		
									A071202	Inaccurate Synchronization		
							E0717	Dyspnea	A0712	Pacing Problem		
									A071202	Inaccurate Synchronization		
							F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A0712	Pacing Problem		
									A071202	Inaccurate Synchronization		
11956	2023	All other reportable incidents	47263	Cardiac resynchronization therapy implantable pacemaker	CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A070101	Failure to Capture		
									A0712	Pacing Problem		
							F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A070101	Failure to Capture		
									A0712	Pacing Problem		
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A070101	Failure to Capture		
									A0712	Pacing Problem		
							F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A070101	Failure to Capture		
									A0712	Pacing Problem		
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A070101	Failure to Capture		
									A0712	Pacing Problem		
							F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A070101	Failure to Capture		
									A0712	Pacing Problem		
11957	2023	All other reportable incidents	47265	Dual-chamber implantable pacemaker, rate-responsive	CT141	Electrodes/Leads and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem		
							F1901	Additional Surgery	A24	No Apparent Device Problem		
					CT145	Cardiovascular devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem		
							F1901	Additional Surgery	A24	No Apparent Device Problem		
					CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem		
							F1901	Additional Surgery	A24	No Apparent Device Problem		
					CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem		
							F1901	Additional Surgery	A24	No Apparent Device Problem		
					CT233	Surgical	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem		
							F1901	Additional Surgery	A24	No Apparent Device Problem		
					CT2406	Implantable	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem		
							F1901	Additional Surgery	A24	No Apparent Device Problem		
					CT2407	Active implantable	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem		
							F1901	Additional Surgery	A24	No Apparent Device Problem		
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem		
							F1901	Additional Surgery	A24	No Apparent Device Problem		
					CT2727	Active implant devices/ancillaries	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem		
							F1901	Additional Surgery	A24	No Apparent Device Problem		
					CT2750	Implantable cardiac electrical stimulators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem		
							F1901	Additional Surgery	A24	No Apparent Device Problem		
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem		
							F1901	Additional Surgery	A24	No Apparent Device Problem		
					CT2985	Active	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem		
							F1901	Additional Surgery	A24	No Apparent Device Problem		
					CT3001	Implantable metal	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem		
							F1901	Additional Surgery	A24	No Apparent Device Problem		

					CT3002	Implantable synthetic polymer	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem		
							F1901	Additional Surgery	A24	No Apparent Device Problem		
					CT3006	Implantable inorganic material	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem		
							F1901	Additional Surgery	A24	No Apparent Device Problem		
					CT314	Electrically powered	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem		
							F1901	Additional Surgery	A24	No Apparent Device Problem		
					CT321	Long-term surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem		
							F1901	Additional Surgery	A24	No Apparent Device Problem		
					CT449	Generators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem		
							F1901	Additional Surgery	A24	No Apparent Device Problem		
					CT981	Single use	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem		
							F1901	Additional Surgery	A24	No Apparent Device Problem		
					CT983	Surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem		
							F1901	Additional Surgery	A24	No Apparent Device Problem		
11958	2023		47265	Dual-chamber implantable pacemaker, rate-responsive	CT141	Electrodes/Leads and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	02/09/2022	02/09/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	02/09/2022	02/09/2022
					CT145	Cardiovascular devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	02/09/2022	02/09/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	02/09/2022	02/09/2022
					CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	02/09/2022	02/09/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	02/09/2022	02/09/2022
					CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	02/09/2022	02/09/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	02/09/2022	02/09/2022
					CT233	Surgical	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	02/09/2022	02/09/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	02/09/2022	02/09/2022
					CT2406	Implantable	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	02/09/2022	02/09/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	02/09/2022	02/09/2022
					CT2407	Active implantable	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	02/09/2022	02/09/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	02/09/2022	02/09/2022
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	02/09/2022	02/09/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	02/09/2022	02/09/2022
					CT2727	Active implant devices/ancillaries	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	02/09/2022	02/09/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	02/09/2022	02/09/2022
					CT2750	Implantable cardiac electrical stimulators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	02/09/2022	02/09/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	02/09/2022	02/09/2022
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	02/09/2022	02/09/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	02/09/2022	02/09/2022
					CT2985	Active	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	02/09/2022	02/09/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	02/09/2022	02/09/2022
					CT3001	Implantable metal	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	02/09/2022	02/09/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	02/09/2022	02/09/2022
					CT3002	Implantable synthetic polymer	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	02/09/2022	02/09/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	02/09/2022	02/09/2022
					CT3006	Implantable inorganic material	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	02/09/2022	02/09/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	02/09/2022	02/09/2022
					CT314	Electrically powered	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	02/09/2022	02/09/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	02/09/2022	02/09/2022
					CT321	Long-term surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	02/09/2022	02/09/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	02/09/2022	02/09/2022
					CT449	Generators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	02/09/2022	02/09/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	02/09/2022	02/09/2022
					CT981	Single use	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	02/09/2022	02/09/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	02/09/2022	02/09/2022
					CT983	Surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	02/09/2022	02/09/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	02/09/2022	02/09/2022
11959	2023		47265	Dual-chamber implantable pacemaker, rate-responsive	CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem		
							F1901	Additional Surgery	A24	No Apparent Device Problem		
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem		
							F1901	Additional Surgery	A24	No Apparent Device Problem		
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem		
							F1901	Additional Surgery	A24	No Apparent Device Problem		
11960	2023		47265	Dual-chamber implantable pacemaker, rate-responsive	CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem		
							F1901	Additional Surgery	A24	No Apparent Device Problem		
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem		
							F1901	Additional Surgery	A24	No Apparent Device Problem		
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem		
							F1901	Additional Surgery	A24	No Apparent Device Problem		
11961	2023		47265	Dual-chamber implantable pacemaker, rate-responsive	CT141	Electrodes/Leads and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	29/09/2022	29/09/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	29/09/2022	29/09/2022
					CT145	Cardiovascular devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	29/09/2022	29/09/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	29/09/2022	29/09/2022
					CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	29/09/2022	29/09/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	29/09/2022	29/09/2022
					CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	29/09/2022	29/09/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	29/09/2022	29/09/2022
					CT233	Surgical	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	29/09/2022	29/09/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	29/09/2022	29/09/2022
					CT2406	Implantable	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	29/09/2022	29/09/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	29/09/2022	29/09/2022
					CT2407	Active implantable	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	29/09/2022	29/09/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	29/09/2022	29/09/2022
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	29/09/2022	29/09/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	29/09/2022	29/09/2022
					CT2727	Active implant devices/ancillaries	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	29/09/2022	29/09/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	29/09/2022	29/09/2022
					CT2750	Implantable cardiac electrical stimulators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	29/09/2022	29/09/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	29/09/2022	29/09/2022
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	29/09/2022	29/09/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	29/09/2022	29/09/2022

					CT2985	Active	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	29/09/2022	29/09/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	29/09/2022	29/09/2022
					CT3001	Implantable metal	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	29/09/2022	29/09/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	29/09/2022	29/09/2022
					CT3002	Implantable synthetic polymer	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	29/09/2022	29/09/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	29/09/2022	29/09/2022
					CT3006	Implantable inorganic material	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	29/09/2022	29/09/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	29/09/2022	29/09/2022
					CT314	Electrically powered	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	29/09/2022	29/09/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	29/09/2022	29/09/2022
					CT321	Long-term surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	29/09/2022	29/09/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	29/09/2022	29/09/2022
					CT449	Generators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	29/09/2022	29/09/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	29/09/2022	29/09/2022
					CT981	Single use	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	29/09/2022	29/09/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	29/09/2022	29/09/2022
					CT983	Surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	29/09/2022	29/09/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	29/09/2022	29/09/2022
11962	2023	All other reportable incidents	47265	Dual-chamber implantable pacemaker, rate-responsive	CT141	Electrodes/Leads and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A071205	Pacemaker Found in Back-Up Mode		
							F1901	Additional Surgery	A071205	Pacemaker Found in Back-Up Mode		
					CT145	Cardiovascular devices	E2403	No Clinical Signs, Symptoms or Conditions	A071205	Pacemaker Found in Back-Up Mode		
							F1901	Additional Surgery	A071205	Pacemaker Found in Back-Up Mode		
					CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A071205	Pacemaker Found in Back-Up Mode		
							F1901	Additional Surgery	A071205	Pacemaker Found in Back-Up Mode		
					CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A071205	Pacemaker Found in Back-Up Mode		
							F1901	Additional Surgery	A071205	Pacemaker Found in Back-Up Mode		
					CT233	Surgical	E2403	No Clinical Signs, Symptoms or Conditions	A071205	Pacemaker Found in Back-Up Mode		
							F1901	Additional Surgery	A071205	Pacemaker Found in Back-Up Mode		
					CT2406	Implantable	E2403	No Clinical Signs, Symptoms or Conditions	A071205	Pacemaker Found in Back-Up Mode		
							F1901	Additional Surgery	A071205	Pacemaker Found in Back-Up Mode		
					CT2407	Active implantable	E2403	No Clinical Signs, Symptoms or Conditions	A071205	Pacemaker Found in Back-Up Mode		
							F1901	Additional Surgery	A071205	Pacemaker Found in Back-Up Mode		
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A071205	Pacemaker Found in Back-Up Mode		
							F1901	Additional Surgery	A071205	Pacemaker Found in Back-Up Mode		
					CT2727	Active implant devices/ancillaries	E2403	No Clinical Signs, Symptoms or Conditions	A071205	Pacemaker Found in Back-Up Mode		
							F1901	Additional Surgery	A071205	Pacemaker Found in Back-Up Mode		
					CT2750	Implantable cardiac electrical stimulators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A071205	Pacemaker Found in Back-Up Mode		
							F1901	Additional Surgery	A071205	Pacemaker Found in Back-Up Mode		
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A071205	Pacemaker Found in Back-Up Mode		
							F1901	Additional Surgery	A071205	Pacemaker Found in Back-Up Mode		
					CT2985	Active	E2403	No Clinical Signs, Symptoms or Conditions	A071205	Pacemaker Found in Back-Up Mode		
							F1901	Additional Surgery	A071205	Pacemaker Found in Back-Up Mode		
					CT3001	Implantable metal	E2403	No Clinical Signs, Symptoms or Conditions	A071205	Pacemaker Found in Back-Up Mode		
							F1901	Additional Surgery	A071205	Pacemaker Found in Back-Up Mode		
					CT3002	Implantable synthetic polymer	E2403	No Clinical Signs, Symptoms or Conditions	A071205	Pacemaker Found in Back-Up Mode		
							F1901	Additional Surgery	A071205	Pacemaker Found in Back-Up Mode		
					CT3006	Implantable inorganic material	E2403	No Clinical Signs, Symptoms or Conditions	A071205	Pacemaker Found in Back-Up Mode		
							F1901	Additional Surgery	A071205	Pacemaker Found in Back-Up Mode		
					CT314	Electrically powered	E2403	No Clinical Signs, Symptoms or Conditions	A071205	Pacemaker Found in Back-Up Mode		
							F1901	Additional Surgery	A071205	Pacemaker Found in Back-Up Mode		
					CT321	Long-term surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A071205	Pacemaker Found in Back-Up Mode		
							F1901	Additional Surgery	A071205	Pacemaker Found in Back-Up Mode		
					CT449	Generators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A071205	Pacemaker Found in Back-Up Mode		
							F1901	Additional Surgery	A071205	Pacemaker Found in Back-Up Mode		
					CT981	Single use	E2403	No Clinical Signs, Symptoms or Conditions	A071205	Pacemaker Found in Back-Up Mode		
							F1901	Additional Surgery	A071205	Pacemaker Found in Back-Up Mode		
					CT983	Surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A071205	Pacemaker Found in Back-Up Mode		
							F1901	Additional Surgery	A071205	Pacemaker Found in Back-Up Mode		
11963	2023	All other reportable incidents	35852	Single-chamber implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E011903	Syncope/Fainting	A070910	Undersensing		
									A0909	Unexpected Therapeutic Results		
							F08	Hospitalization or Prolonged Hospitalization	A070910	Undersensing		
									A0909	Unexpected Therapeutic Results		
11964	2023	All other reportable incidents	47265	Dual-chamber implantable pacemaker, rate-responsive	CT141	Electrodes/Leads and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem		
									A150103	Premature Activation		
							F26	No Health Consequences or Impact	A0705	Battery Problem		
									A150103	Premature Activation		
					CT145	Cardiovascular devices	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem		
									A150103	Premature Activation		
							F26	No Health Consequences or Impact	A0705	Battery Problem		
									A150103	Premature Activation		
					CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem		
									A150103	Premature Activation		
							F26	No Health Consequences or Impact	A0705	Battery Problem		
									A150103	Premature Activation		
					CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem		
									A150103	Premature Activation		
							F26	No Health Consequences or Impact	A0705	Battery Problem		
									A150103	Premature Activation		
					CT233	Surgical	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem		
									A150103	Premature Activation		
							F26	No Health Consequences or Impact	A0705	Battery Problem		
									A150103	Premature Activation		
					CT2406	Implantable	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem		
									A150103	Premature Activation		
							F26	No Health Consequences or Impact	A0705	Battery Problem		
									A150103	Premature Activation		
					CT2407	Active implantable	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem		
									A150103	Premature Activation		

							F26	No Health Consequences or Impact	A0705	Battery Problem		
									A150103	Premature Activation		
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem		
									A150103	Premature Activation		
							F26	No Health Consequences or Impact	A0705	Battery Problem		
									A150103	Premature Activation		
					CT2727	Active implant devices/ancillaries	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem		
									A150103	Premature Activation		
							F26	No Health Consequences or Impact	A0705	Battery Problem		
									A150103	Premature Activation		
					CT2750	Implantable cardiac electrical stimulators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem		
									A150103	Premature Activation		
							F26	No Health Consequences or Impact	A0705	Battery Problem		
									A150103	Premature Activation		
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem		
									A150103	Premature Activation		
							F26	No Health Consequences or Impact	A0705	Battery Problem		
									A150103	Premature Activation		
					CT2985	Active	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem		
									A150103	Premature Activation		
							F26	No Health Consequences or Impact	A0705	Battery Problem		
									A150103	Premature Activation		
					CT3001	Implantable metal	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem		
									A150103	Premature Activation		
							F26	No Health Consequences or Impact	A0705	Battery Problem		
									A150103	Premature Activation		
					CT3002	Implantable synthetic polymer	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem		
									A150103	Premature Activation		
							F26	No Health Consequences or Impact	A0705	Battery Problem		
									A150103	Premature Activation		
					CT3006	Implantable inorganic material	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem		
									A150103	Premature Activation		
							F26	No Health Consequences or Impact	A0705	Battery Problem		
									A150103	Premature Activation		
					CT314	Electrically powered	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem		
									A150103	Premature Activation		
							F26	No Health Consequences or Impact	A0705	Battery Problem		
									A150103	Premature Activation		
					CT321	Long-term surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem		
									A150103	Premature Activation		
							F26	No Health Consequences or Impact	A0705	Battery Problem		
									A150103	Premature Activation		
					CT449	Generators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem		
									A150103	Premature Activation		
							F26	No Health Consequences or Impact	A0705	Battery Problem		
									A150103	Premature Activation		
					CT981	Single use	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem		
									A150103	Premature Activation		
							F26	No Health Consequences or Impact	A0705	Battery Problem		
									A150103	Premature Activation		
					CT983	Surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem		
									A150103	Premature Activation		
							F26	No Health Consequences or Impact	A0705	Battery Problem		
									A150103	Premature Activation		
11965	2023		47265	Dual-chamber implantable pacemaker, rate-responsive	CT141	Electrodes/Leads and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem		
							F1901	Additional Surgery	A24	No Apparent Device Problem		
					CT145	Cardiovascular devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem		
							F1901	Additional Surgery	A24	No Apparent Device Problem		
					CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem		
							F1901	Additional Surgery	A24	No Apparent Device Problem		
					CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem		
							F1901	Additional Surgery	A24	No Apparent Device Problem		
					CT233	Surgical	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem		
							F1901	Additional Surgery	A24	No Apparent Device Problem		
					CT2406	Implantable	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem		
							F1901	Additional Surgery	A24	No Apparent Device Problem		
					CT2407	Active implantable	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem		
							F1901	Additional Surgery	A24	No Apparent Device Problem		
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem		
							F1901	Additional Surgery	A24	No Apparent Device Problem		
					CT2727	Active implant devices/ancillaries	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem		
							F1901	Additional Surgery	A24	No Apparent Device Problem		
					CT2750	Implantable cardiac electrical stimulators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem		
							F1901	Additional Surgery	A24	No Apparent Device Problem		
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem		
							F1901	Additional Surgery	A24	No Apparent Device Problem		
					CT2985	Active	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem		
							F1901	Additional Surgery	A24	No Apparent Device Problem		
					CT3001	Implantable metal	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem		
							F1901	Additional Surgery	A24	No Apparent Device Problem		
					CT3002	Implantable synthetic polymer	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem		
							F1901	Additional Surgery	A24	No Apparent Device Problem		
					CT3006	Implantable inorganic material	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem		
							F1901	Additional Surgery	A24	No Apparent Device Problem		
					CT314	Electrically powered	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem		
							F1901	Additional Surgery	A24	No Apparent Device Problem		
					CT321	Long-term surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem		
							F1901	Additional Surgery	A24	No Apparent Device Problem		

					CT449	Generators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem		
							F1901	Additional Surgery	A24	No Apparent Device Problem		
					CT981	Single use	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem		
							F1901	Additional Surgery	A24	No Apparent Device Problem		
					CT983	Surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem		
							F1901	Additional Surgery	A24	No Apparent Device Problem		
11966	2023	All other reportable incidents	47265	Dual-chamber implantable pacemaker, rate-responsive	CT141	Electrodes/Leads and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A071102	Failure to Interrogate	30/12/2022	30/12/2022
									A0712	Pacing Problem	30/12/2022	30/12/2022
							F1901	Additional Surgery	A071102	Failure to Interrogate	30/12/2022	30/12/2022
									A0712	Pacing Problem	30/12/2022	30/12/2022
					CT145	Cardiovascular devices	E2403	No Clinical Signs, Symptoms or Conditions	A071102	Failure to Interrogate	30/12/2022	30/12/2022
									A0712	Pacing Problem	30/12/2022	30/12/2022
							F1901	Additional Surgery	A071102	Failure to Interrogate	30/12/2022	30/12/2022
									A0712	Pacing Problem	30/12/2022	30/12/2022
					CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A071102	Failure to Interrogate	30/12/2022	30/12/2022
									A0712	Pacing Problem	30/12/2022	30/12/2022
							F1901	Additional Surgery	A071102	Failure to Interrogate	30/12/2022	30/12/2022
									A0712	Pacing Problem	30/12/2022	30/12/2022
					CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A071102	Failure to Interrogate	30/12/2022	30/12/2022
									A0712	Pacing Problem	30/12/2022	30/12/2022
							F1901	Additional Surgery	A071102	Failure to Interrogate	30/12/2022	30/12/2022
									A0712	Pacing Problem	30/12/2022	30/12/2022
					CT233	Surgical	E2403	No Clinical Signs, Symptoms or Conditions	A071102	Failure to Interrogate	30/12/2022	30/12/2022
									A0712	Pacing Problem	30/12/2022	30/12/2022
							F1901	Additional Surgery	A071102	Failure to Interrogate	30/12/2022	30/12/2022
									A0712	Pacing Problem	30/12/2022	30/12/2022
					CT2406	Implantable	E2403	No Clinical Signs, Symptoms or Conditions	A071102	Failure to Interrogate	30/12/2022	30/12/2022
									A0712	Pacing Problem	30/12/2022	30/12/2022
							F1901	Additional Surgery	A071102	Failure to Interrogate	30/12/2022	30/12/2022
									A0712	Pacing Problem	30/12/2022	30/12/2022
					CT2407	Active implantable	E2403	No Clinical Signs, Symptoms or Conditions	A071102	Failure to Interrogate	30/12/2022	30/12/2022
									A0712	Pacing Problem	30/12/2022	30/12/2022
							F1901	Additional Surgery	A071102	Failure to Interrogate	30/12/2022	30/12/2022
									A0712	Pacing Problem	30/12/2022	30/12/2022
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A071102	Failure to Interrogate	30/12/2022	30/12/2022
									A0712	Pacing Problem	30/12/2022	30/12/2022
							F1901	Additional Surgery	A071102	Failure to Interrogate	30/12/2022	30/12/2022
									A0712	Pacing Problem	30/12/2022	30/12/2022
					CT2727	Active implant devices/ancillaries	E2403	No Clinical Signs, Symptoms or Conditions	A071102	Failure to Interrogate	30/12/2022	30/12/2022
									A0712	Pacing Problem	30/12/2022	30/12/2022
							F1901	Additional Surgery	A071102	Failure to Interrogate	30/12/2022	30/12/2022
									A0712	Pacing Problem	30/12/2022	30/12/2022
					CT2750	Implantable cardiac electrical stimulators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A071102	Failure to Interrogate	30/12/2022	30/12/2022
									A0712	Pacing Problem	30/12/2022	30/12/2022
							F1901	Additional Surgery	A071102	Failure to Interrogate	30/12/2022	30/12/2022
									A0712	Pacing Problem	30/12/2022	30/12/2022
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A071102	Failure to Interrogate	30/12/2022	30/12/2022
									A0712	Pacing Problem	30/12/2022	30/12/2022
							F1901	Additional Surgery	A071102	Failure to Interrogate	30/12/2022	30/12/2022
									A0712	Pacing Problem	30/12/2022	30/12/2022
					CT2985	Active	E2403	No Clinical Signs, Symptoms or Conditions	A071102	Failure to Interrogate	30/12/2022	30/12/2022
									A0712	Pacing Problem	30/12/2022	30/12/2022
							F1901	Additional Surgery	A071102	Failure to Interrogate	30/12/2022	30/12/2022
									A0712	Pacing Problem	30/12/2022	30/12/2022
					CT3001	Implantable metal	E2403	No Clinical Signs, Symptoms or Conditions	A071102	Failure to Interrogate	30/12/2022	30/12/2022
									A0712	Pacing Problem	30/12/2022	30/12/2022
							F1901	Additional Surgery	A071102	Failure to Interrogate	30/12/2022	30/12/2022
									A0712	Pacing Problem	30/12/2022	30/12/2022
					CT3002	Implantable synthetic polymer	E2403	No Clinical Signs, Symptoms or Conditions	A071102	Failure to Interrogate	30/12/2022	30/12/2022
									A0712	Pacing Problem	30/12/2022	30/12/2022
							F1901	Additional Surgery	A071102	Failure to Interrogate	30/12/2022	30/12/2022
									A0712	Pacing Problem	30/12/2022	30/12/2022
					CT3006	Implantable inorganic material	E2403	No Clinical Signs, Symptoms or Conditions	A071102	Failure to Interrogate	30/12/2022	30/12/2022
									A0712	Pacing Problem	30/12/2022	30/12/2022
							F1901	Additional Surgery	A071102	Failure to Interrogate	30/12/2022	30/12/2022
									A0712	Pacing Problem	30/12/2022	30/12/2022
					CT314	Electrically powered	E2403	No Clinical Signs, Symptoms or Conditions	A071102	Failure to Interrogate	30/12/2022	30/12/2022
									A0712	Pacing Problem	30/12/2022	30/12/2022
							F1901	Additional Surgery	A071102	Failure to Interrogate	30/12/2022	30/12/2022
									A0712	Pacing Problem	30/12/2022	30/12/2022
					CT321	Long-term surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A071102	Failure to Interrogate	30/12/2022	30/12/2022
									A0712	Pacing Problem	30/12/2022	30/12/2022
							F1901	Additional Surgery	A071102	Failure to Interrogate	30/12/2022	30/12/2022
									A0712	Pacing Problem	30/12/2022	30/12/2022
					CT449	Generators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A071102	Failure to Interrogate	30/12/2022	30/12/2022
									A0712	Pacing Problem	30/12/2022	30/12/2022
							F1901	Additional Surgery	A071102	Failure to Interrogate	30/12/2022	30/12/2022
									A0712	Pacing Problem	30/12/2022	30/12/2022
					CT981	Single use	E2403	No Clinical Signs, Symptoms or Conditions	A071102	Failure to Interrogate	30/12/2022	30/12/2022
									A0712	Pacing Problem	30/12/2022	30/12/2022
							F1901	Additional Surgery	A071102	Failure to Interrogate	30/12/2022	30/12/2022
									A0712	Pacing Problem	30/12/2022	30/12/2022
					CT983	Surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A071102	Failure to Interrogate	30/12/2022	30/12/2022
									A0712	Pacing Problem	30/12/2022	30/12/2022
							F1901	Additional Surgery	A071102	Failure to Interrogate	30/12/2022	30/12/2022
									A0712	Pacing Problem	30/12/2022	30/12/2022
11967	2023	All other reportable incidents	47265	Dual-chamber implantable pacemaker, rate-responsive	CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem	24/10/2022	24/10/2022
									A160601	Premature Elective Replacement Indicator	24/10/2022	24/10/2022

							F08	Hospitalization or Prolonged Hospitalization	A0705	Battery Problem	24/10/2022	24/10/2022
									A160601	Premature Elective Replacement Indicator	24/10/2022	24/10/2022
							F1905	Device Revision or Replacement	A0705	Battery Problem	24/10/2022	24/10/2022
									A160601	Premature Elective Replacement Indicator	24/10/2022	24/10/2022
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem	24/10/2022	24/10/2022
									A160601	Premature Elective Replacement Indicator	24/10/2022	24/10/2022
							F08	Hospitalization or Prolonged Hospitalization	A0705	Battery Problem	24/10/2022	24/10/2022
									A160601	Premature Elective Replacement Indicator	24/10/2022	24/10/2022
							F1905	Device Revision or Replacement	A0705	Battery Problem	24/10/2022	24/10/2022
									A160601	Premature Elective Replacement Indicator	24/10/2022	24/10/2022
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem	24/10/2022	24/10/2022
									A160601	Premature Elective Replacement Indicator	24/10/2022	24/10/2022
							F08	Hospitalization or Prolonged Hospitalization	A0705	Battery Problem	24/10/2022	24/10/2022
									A160601	Premature Elective Replacement Indicator	24/10/2022	24/10/2022
							F1905	Device Revision or Replacement	A0705	Battery Problem	24/10/2022	24/10/2022
									A160601	Premature Elective Replacement Indicator	24/10/2022	24/10/2022
11968	2023	All other reportable incidents	47263	Cardiac resynchronization therapy implantable pacemaker	CT1684	Cardiac pacemakers and associated devices	E0601	Arrhythmia	A070101	Failure to Capture		
							F23	Unexpected Medical Intervention	A070101	Failure to Capture		
					CT1685	Cardiac pacemakers	E0601	Arrhythmia	A070101	Failure to Capture		
							F23	Unexpected Medical Intervention	A070101	Failure to Capture		
					CT257	Implantable cardiac pacemakers	E0601	Arrhythmia	A070101	Failure to Capture		
							F23	Unexpected Medical Intervention	A070101	Failure to Capture		
					CT2866	Implantable cardiac electrical stimulators	E0601	Arrhythmia	A070101	Failure to Capture		
							F23	Unexpected Medical Intervention	A070101	Failure to Capture		
11969	2023	All other reportable incidents	47265	Dual-chamber implantable pacemaker, rate-responsive	CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A070910	Undersensing		
							F15	Recognized Device or Procedural Complication	A070910	Undersensing		
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A070910	Undersensing		
							F15	Recognized Device or Procedural Complication	A070910	Undersensing		
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A070910	Undersensing		
							F15	Recognized Device or Procedural Complication	A070910	Undersensing		
11970	2023		61439	Cardiac pulse generator reader, professional	CT1684	Cardiac pacemakers and associated devices	E23	Generalized Disorders	A0713	Defibrillation/Stimulation Problem		
							F05	Delay to Treatment/ Therapy	A0713	Defibrillation/Stimulation Problem		
							F25	Unanticipated Adverse Device Effect	A0713	Defibrillation/Stimulation Problem		
11971	2023		47265	Dual-chamber implantable pacemaker, rate-responsive	CT141	Electrodes/Leads and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	25/11/2022	25/11/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	25/11/2022	25/11/2022
					CT145	Cardiovascular devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	25/11/2022	25/11/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	25/11/2022	25/11/2022
					CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	25/11/2022	25/11/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	25/11/2022	25/11/2022
					CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	25/11/2022	25/11/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	25/11/2022	25/11/2022
					CT233	Surgical	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	25/11/2022	25/11/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	25/11/2022	25/11/2022
					CT2406	Implantable	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	25/11/2022	25/11/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	25/11/2022	25/11/2022
					CT2407	Active implantable	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	25/11/2022	25/11/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	25/11/2022	25/11/2022
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	25/11/2022	25/11/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	25/11/2022	25/11/2022
					CT2727	Active implant devices/ancillaries	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	25/11/2022	25/11/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	25/11/2022	25/11/2022
					CT2750	Implantable cardiac electrical stimulators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	25/11/2022	25/11/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	25/11/2022	25/11/2022
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	25/11/2022	25/11/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	25/11/2022	25/11/2022
					CT2985	Active	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	25/11/2022	25/11/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	25/11/2022	25/11/2022
					CT3001	Implantable metal	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	25/11/2022	25/11/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	25/11/2022	25/11/2022
					CT3002	Implantable synthetic polymer	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	25/11/2022	25/11/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	25/11/2022	25/11/2022
					CT3006	Implantable inorganic material	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	25/11/2022	25/11/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	25/11/2022	25/11/2022
					CT314	Electrically powered	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	25/11/2022	25/11/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	25/11/2022	25/11/2022
					CT321	Long-term surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	25/11/2022	25/11/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	25/11/2022	25/11/2022
					CT449	Generators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	25/11/2022	25/11/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	25/11/2022	25/11/2022
					CT981	Single use	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	25/11/2022	25/11/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	25/11/2022	25/11/2022
					CT983	Surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	25/11/2022	25/11/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	25/11/2022	25/11/2022
11972	2023		47265	Dual-chamber implantable pacemaker, rate-responsive	CT141	Electrodes/Leads and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	07/12/2022	07/12/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	07/12/2022	07/12/2022
					CT145	Cardiovascular devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	07/12/2022	07/12/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	07/12/2022	07/12/2022
					CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	07/12/2022	07/12/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	07/12/2022	07/12/2022
					CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	07/12/2022	07/12/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	07/12/2022	07/12/2022
					CT233	Surgical	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	07/12/2022	07/12/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	07/12/2022	07/12/2022
					CT2406	Implantable	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	07/12/2022	07/12/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	07/12/2022	07/12/2022
					CT2407	Active implantable	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	07/12/2022	07/12/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	07/12/2022	07/12/2022
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	07/12/2022	07/12/2022

						F1901	Additional Surgery	A24	No Apparent Device Problem	07/12/2022	07/12/2022
					CT2727	Active implant devices/ancillaries	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	07/12/2022	07/12/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	07/12/2022	07/12/2022
					CT2750	Implantable cardiac electrical stimulators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	07/12/2022	07/12/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	07/12/2022	07/12/2022
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	07/12/2022	07/12/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	07/12/2022	07/12/2022
					CT2985	Active	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	07/12/2022	07/12/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	07/12/2022	07/12/2022
					CT3001	Implantable metal	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	07/12/2022	07/12/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	07/12/2022	07/12/2022
					CT3002	Implantable synthetic polymer	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	07/12/2022	07/12/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	07/12/2022	07/12/2022
					CT3006	Implantable inorganic material	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	07/12/2022	07/12/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	07/12/2022	07/12/2022
					CT314	Electrically powered	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	07/12/2022	07/12/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	07/12/2022	07/12/2022
					CT321	Long-term surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	07/12/2022	07/12/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	07/12/2022	07/12/2022
					CT449	Generators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	07/12/2022	07/12/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	07/12/2022	07/12/2022
					CT981	Single use	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	07/12/2022	07/12/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	07/12/2022	07/12/2022
					CT983	Surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	07/12/2022	07/12/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	07/12/2022	07/12/2022
11973	2023		47265	Dual-chamber implantable pacemaker, rate-responsive	CT141	Electrodes/Leads and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	12/01/2023	12/01/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	12/01/2023	12/01/2023
					CT145	Cardiovascular devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	12/01/2023	12/01/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	12/01/2023	12/01/2023
					CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	12/01/2023	12/01/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	12/01/2023	12/01/2023
					CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	12/01/2023	12/01/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	12/01/2023	12/01/2023
					CT233	Surgical	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	12/01/2023	12/01/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	12/01/2023	12/01/2023
					CT2406	Implantable	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	12/01/2023	12/01/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	12/01/2023	12/01/2023
					CT2407	Active implantable	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	12/01/2023	12/01/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	12/01/2023	12/01/2023
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	12/01/2023	12/01/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	12/01/2023	12/01/2023
					CT2727	Active implant devices/ancillaries	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	12/01/2023	12/01/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	12/01/2023	12/01/2023
					CT2750	Implantable cardiac electrical stimulators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	12/01/2023	12/01/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	12/01/2023	12/01/2023
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	12/01/2023	12/01/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	12/01/2023	12/01/2023
					CT2985	Active	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	12/01/2023	12/01/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	12/01/2023	12/01/2023
					CT3001	Implantable metal	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	12/01/2023	12/01/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	12/01/2023	12/01/2023
					CT3002	Implantable synthetic polymer	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	12/01/2023	12/01/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	12/01/2023	12/01/2023
					CT3006	Implantable inorganic material	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	12/01/2023	12/01/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	12/01/2023	12/01/2023
					CT314	Electrically powered	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	12/01/2023	12/01/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	12/01/2023	12/01/2023
					CT321	Long-term surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	12/01/2023	12/01/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	12/01/2023	12/01/2023
					CT449	Generators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	12/01/2023	12/01/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	12/01/2023	12/01/2023
					CT981	Single use	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	12/01/2023	12/01/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	12/01/2023	12/01/2023
					CT983	Surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	12/01/2023	12/01/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	12/01/2023	12/01/2023
11974	2023		47265	Dual-chamber implantable pacemaker, rate-responsive	CT141	Electrodes/Leads and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	07/01/2023	07/01/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	07/01/2023	07/01/2023
					CT145	Cardiovascular devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	07/01/2023	07/01/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	07/01/2023	07/01/2023
					CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	07/01/2023	07/01/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	07/01/2023	07/01/2023
					CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	07/01/2023	07/01/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	07/01/2023	07/01/2023
					CT233	Surgical	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	07/01/2023	07/01/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	07/01/2023	07/01/2023
					CT2406	Implantable	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	07/01/2023	07/01/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	07/01/2023	07/01/2023
					CT2407	Active implantable	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	07/01/2023	07/01/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	07/01/2023	07/01/2023
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	07/01/2023	07/01/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	07/01/2023	07/01/2023
					CT2727	Active implant devices/ancillaries	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	07/01/2023	07/01/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	07/01/2023	07/01/2023
					CT2750	Implantable cardiac electrical stimulators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	07/01/2023	07/01/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	07/01/2023	07/01/2023
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	07/01/2023	07/01/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	07/01/2023	07/01/2023
					CT2985	Active	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	07/01/2023	07/01/2023

							F1901	Additional Surgery	A24	No Apparent Device Problem	07/01/2023	07/01/2023
					CT3001	Implantable metal	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	07/01/2023	07/01/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	07/01/2023	07/01/2023
					CT3002	Implantable synthetic polymer	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	07/01/2023	07/01/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	07/01/2023	07/01/2023
					CT3006	Implantable inorganic material	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	07/01/2023	07/01/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	07/01/2023	07/01/2023
					CT314	Electrically powered	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	07/01/2023	07/01/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	07/01/2023	07/01/2023
					CT321	Long-term surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	07/01/2023	07/01/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	07/01/2023	07/01/2023
					CT449	Generators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	07/01/2023	07/01/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	07/01/2023	07/01/2023
					CT981	Single use	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	07/01/2023	07/01/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	07/01/2023	07/01/2023
					CT983	Surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	07/01/2023	07/01/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	07/01/2023	07/01/2023
11975	2023		47265	Dual-chamber implantable pacemaker, rate-responsive	CT141	Electrodes/Leads and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	03/01/2023	03/01/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	03/01/2023	03/01/2023
					CT145	Cardiovascular devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	03/01/2023	03/01/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	03/01/2023	03/01/2023
					CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	03/01/2023	03/01/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	03/01/2023	03/01/2023
					CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	03/01/2023	03/01/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	03/01/2023	03/01/2023
					CT233	Surgical	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	03/01/2023	03/01/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	03/01/2023	03/01/2023
					CT2406	Implantable	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	03/01/2023	03/01/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	03/01/2023	03/01/2023
					CT2407	Active implantable	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	03/01/2023	03/01/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	03/01/2023	03/01/2023
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	03/01/2023	03/01/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	03/01/2023	03/01/2023
					CT2727	Active implant devices/ancillaries	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	03/01/2023	03/01/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	03/01/2023	03/01/2023
					CT2750	Implantable cardiac electrical stimulators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	03/01/2023	03/01/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	03/01/2023	03/01/2023
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	03/01/2023	03/01/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	03/01/2023	03/01/2023
					CT2985	Active	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	03/01/2023	03/01/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	03/01/2023	03/01/2023
					CT3001	Implantable metal	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	03/01/2023	03/01/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	03/01/2023	03/01/2023
					CT3002	Implantable synthetic polymer	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	03/01/2023	03/01/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	03/01/2023	03/01/2023
					CT3006	Implantable inorganic material	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	03/01/2023	03/01/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	03/01/2023	03/01/2023
					CT314	Electrically powered	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	03/01/2023	03/01/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	03/01/2023	03/01/2023
					CT321	Long-term surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	03/01/2023	03/01/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	03/01/2023	03/01/2023
					CT449	Generators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	03/01/2023	03/01/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	03/01/2023	03/01/2023
					CT981	Single use	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	03/01/2023	03/01/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	03/01/2023	03/01/2023
					CT983	Surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	03/01/2023	03/01/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	03/01/2023	03/01/2023
11976	2023		47265	Dual-chamber implantable pacemaker, rate-responsive	CT141	Electrodes/Leads and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	22/12/2022	22/12/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	22/12/2022	22/12/2022
					CT145	Cardiovascular devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	22/12/2022	22/12/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	22/12/2022	22/12/2022
					CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	22/12/2022	22/12/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	22/12/2022	22/12/2022
					CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	22/12/2022	22/12/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	22/12/2022	22/12/2022
					CT233	Surgical	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	22/12/2022	22/12/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	22/12/2022	22/12/2022
					CT2406	Implantable	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	22/12/2022	22/12/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	22/12/2022	22/12/2022
					CT2407	Active implantable	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	22/12/2022	22/12/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	22/12/2022	22/12/2022
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	22/12/2022	22/12/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	22/12/2022	22/12/2022
					CT2727	Active implant devices/ancillaries	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	22/12/2022	22/12/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	22/12/2022	22/12/2022
					CT2750	Implantable cardiac electrical stimulators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	22/12/2022	22/12/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	22/12/2022	22/12/2022
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	22/12/2022	22/12/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	22/12/2022	22/12/2022
					CT2985	Active	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	22/12/2022	22/12/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	22/12/2022	22/12/2022
					CT3001	Implantable metal	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	22/12/2022	22/12/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	22/12/2022	22/12/2022
					CT3002	Implantable synthetic polymer	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	22/12/2022	22/12/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	22/12/2022	22/12/2022
					CT3006	Implantable inorganic material	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	22/12/2022	22/12/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	22/12/2022	22/12/2022
					CT314	Electrically powered	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	22/12/2022	22/12/2022

							F1901	Additional Surgery	A24	No Apparent Device Problem	22/12/2022	22/12/2022
					CT321	Long-term surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	22/12/2022	22/12/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	22/12/2022	22/12/2022
					CT449	Generators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	22/12/2022	22/12/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	22/12/2022	22/12/2022
					CT981	Single use	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	22/12/2022	22/12/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	22/12/2022	22/12/2022
					CT983	Surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	22/12/2022	22/12/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	22/12/2022	22/12/2022
11977	2023	All other reportable incidents	47263	Cardiac resynchronization therapy implantable pacemaker	CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A072201	High Impedance		
							F26	No Health Consequences or Impact	A072201	High Impedance		
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A072201	High Impedance		
							F26	No Health Consequences or Impact	A072201	High Impedance		
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A072201	High Impedance		
							F26	No Health Consequences or Impact	A072201	High Impedance		
11978	2023	All other reportable incidents	17882	Physiologic-monitoring defibrillation system	CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A0719	Unexpected Shutdown		
							F27	Problem Identified During Non-Clinical Procedure	A0719	Unexpected Shutdown		
11979	2023	All other reportable incidents	47265	Dual-chamber implantable pacemaker, rate-responsive	CT141	Electrodes/Leads and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	05/01/2023	05/01/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	05/01/2023	05/01/2023
					CT145	Cardiovascular devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	05/01/2023	05/01/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	05/01/2023	05/01/2023
					CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	05/01/2023	05/01/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	05/01/2023	05/01/2023
					CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	05/01/2023	05/01/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	05/01/2023	05/01/2023
					CT233	Surgical	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	05/01/2023	05/01/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	05/01/2023	05/01/2023
					CT2406	Implantable	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	05/01/2023	05/01/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	05/01/2023	05/01/2023
					CT2407	Active implantable	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	05/01/2023	05/01/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	05/01/2023	05/01/2023
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	05/01/2023	05/01/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	05/01/2023	05/01/2023
					CT2727	Active implant devices/ancillaries	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	05/01/2023	05/01/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	05/01/2023	05/01/2023
					CT2750	Implantable cardiac electrical stimulators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	05/01/2023	05/01/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	05/01/2023	05/01/2023
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	05/01/2023	05/01/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	05/01/2023	05/01/2023
					CT2985	Active	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	05/01/2023	05/01/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	05/01/2023	05/01/2023
					CT3001	Implantable metal	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	05/01/2023	05/01/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	05/01/2023	05/01/2023
					CT3002	Implantable synthetic polymer	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	05/01/2023	05/01/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	05/01/2023	05/01/2023
					CT3006	Implantable inorganic material	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	05/01/2023	05/01/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	05/01/2023	05/01/2023
					CT314	Electrically powered	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	05/01/2023	05/01/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	05/01/2023	05/01/2023
					CT321	Long-term surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	05/01/2023	05/01/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	05/01/2023	05/01/2023
					CT449	Generators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	05/01/2023	05/01/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	05/01/2023	05/01/2023
					CT981	Single use	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	05/01/2023	05/01/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	05/01/2023	05/01/2023
					CT983	Surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	05/01/2023	05/01/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	05/01/2023	05/01/2023
11980	2023		35852	Single-chamber implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	F26	No Health Consequences or Impact	A07	Electrical /Electronic Property Problem		
									A070504	Premature Discharge of Battery		
11981	2023		35852	Single-chamber implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	F26	No Health Consequences or Impact	A0705	Battery Problem		
11982	2023		47265	Dual-chamber implantable pacemaker, rate-responsive	CT141	Electrodes/Leads and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	09/01/2023	09/01/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	09/01/2023	09/01/2023
					CT145	Cardiovascular devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	09/01/2023	09/01/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	09/01/2023	09/01/2023
					CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	09/01/2023	09/01/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	09/01/2023	09/01/2023
					CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	09/01/2023	09/01/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	09/01/2023	09/01/2023
					CT233	Surgical	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	09/01/2023	09/01/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	09/01/2023	09/01/2023
					CT2406	Implantable	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	09/01/2023	09/01/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	09/01/2023	09/01/2023
					CT2407	Active implantable	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	09/01/2023	09/01/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	09/01/2023	09/01/2023
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	09/01/2023	09/01/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	09/01/2023	09/01/2023
					CT2727	Active implant devices/ancillaries	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	09/01/2023	09/01/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	09/01/2023	09/01/2023
					CT2750	Implantable cardiac electrical stimulators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	09/01/2023	09/01/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	09/01/2023	09/01/2023
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	09/01/2023	09/01/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	09/01/2023	09/01/2023
					CT2985	Active	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	09/01/2023	09/01/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	09/01/2023	09/01/2023
					CT3001	Implantable metal	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	09/01/2023	09/01/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	09/01/2023	09/01/2023
					CT3002	Implantable synthetic polymer	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	09/01/2023	09/01/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	09/01/2023	09/01/2023

					CT3006	Implantable inorganic material	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	09/01/2023	09/01/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	09/01/2023	09/01/2023
					CT314	Electrically powered	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	09/01/2023	09/01/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	09/01/2023	09/01/2023
					CT321	Long-term surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	09/01/2023	09/01/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	09/01/2023	09/01/2023
					CT449	Generators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	09/01/2023	09/01/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	09/01/2023	09/01/2023
					CT981	Single use	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	09/01/2023	09/01/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	09/01/2023	09/01/2023
					CT983	Surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	09/01/2023	09/01/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	09/01/2023	09/01/2023
11983	2023		47265	Dual-chamber implantable pacemaker, rate-responsive	CT141	Electrodes/Leads and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	08/10/2022	08/10/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	08/10/2022	08/10/2022
					CT145	Cardiovascular devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	08/10/2022	08/10/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	08/10/2022	08/10/2022
					CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	08/10/2022	08/10/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	08/10/2022	08/10/2022
					CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	08/10/2022	08/10/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	08/10/2022	08/10/2022
					CT233	Surgical	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	08/10/2022	08/10/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	08/10/2022	08/10/2022
					CT2406	Implantable	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	08/10/2022	08/10/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	08/10/2022	08/10/2022
					CT2407	Active implantable	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	08/10/2022	08/10/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	08/10/2022	08/10/2022
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	08/10/2022	08/10/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	08/10/2022	08/10/2022
					CT2727	Active implant devices/ancillaries	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	08/10/2022	08/10/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	08/10/2022	08/10/2022
					CT2750	Implantable cardiac electrical stimulators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	08/10/2022	08/10/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	08/10/2022	08/10/2022
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	08/10/2022	08/10/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	08/10/2022	08/10/2022
					CT2985	Active	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	08/10/2022	08/10/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	08/10/2022	08/10/2022
					CT3001	Implantable metal	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	08/10/2022	08/10/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	08/10/2022	08/10/2022
					CT3002	Implantable synthetic polymer	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	08/10/2022	08/10/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	08/10/2022	08/10/2022
					CT3006	Implantable inorganic material	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	08/10/2022	08/10/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	08/10/2022	08/10/2022
					CT314	Electrically powered	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	08/10/2022	08/10/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	08/10/2022	08/10/2022
					CT321	Long-term surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	08/10/2022	08/10/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	08/10/2022	08/10/2022
					CT449	Generators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	08/10/2022	08/10/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	08/10/2022	08/10/2022
					CT981	Single use	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	08/10/2022	08/10/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	08/10/2022	08/10/2022
					CT983	Surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	08/10/2022	08/10/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	08/10/2022	08/10/2022
11984	2023	All other reportable incidents	47265	Dual-chamber implantable pacemaker, rate-responsive	CT141	Electrodes/Leads and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT145	Cardiovascular devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT233	Surgical	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT2406	Implantable	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT2407	Active implantable	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT2727	Active implant devices/ancillaries	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT2750	Implantable cardiac electrical stimulators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT2985	Active	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT3001	Implantable metal	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT3002	Implantable synthetic polymer	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT3006	Implantable inorganic material	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT314	Electrically powered	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT321	Long-term surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT449	Generators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		

					CT981	Single use	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT983	Surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
11985	2023	All other reportable incidents	47265	Dual-chamber implantable pacemaker, rate-responsive	CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
11986	2023		47265	Dual-chamber implantable pacemaker, rate-responsive	CT141	Electrodes/Leads and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem		
							F02	Death	A24	No Apparent Device Problem		
					CT145	Cardiovascular devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem		
							F02	Death	A24	No Apparent Device Problem		
					CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem		
							F02	Death	A24	No Apparent Device Problem		
					CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem		
							F02	Death	A24	No Apparent Device Problem		
					CT233	Surgical	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem		
							F02	Death	A24	No Apparent Device Problem		
					CT2406	Implantable	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem		
							F02	Death	A24	No Apparent Device Problem		
					CT2407	Active implantable	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem		
							F02	Death	A24	No Apparent Device Problem		
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem		
							F02	Death	A24	No Apparent Device Problem		
					CT2727	Active implant devices/ancillaries	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem		
							F02	Death	A24	No Apparent Device Problem		
					CT2750	Implantable cardiac electrical stimulators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem		
							F02	Death	A24	No Apparent Device Problem		
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem		
							F02	Death	A24	No Apparent Device Problem		
					CT2985	Active	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem		
							F02	Death	A24	No Apparent Device Problem		
					CT3001	Implantable metal	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem		
							F02	Death	A24	No Apparent Device Problem		
					CT3002	Implantable synthetic polymer	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem		
							F02	Death	A24	No Apparent Device Problem		
					CT3006	Implantable inorganic material	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem		
							F02	Death	A24	No Apparent Device Problem		
					CT314	Electrically powered	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem		
							F02	Death	A24	No Apparent Device Problem		
					CT321	Long-term surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem		
							F02	Death	A24	No Apparent Device Problem		
					CT449	Generators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem		
							F02	Death	A24	No Apparent Device Problem		
					CT981	Single use	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem		
							F02	Death	A24	No Apparent Device Problem		
					CT983	Surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem		
							F02	Death	A24	No Apparent Device Problem		
11987	2023		47265	Dual-chamber implantable pacemaker, rate-responsive	CT141	Electrodes/Leads and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	08/10/2022	08/10/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	08/10/2022	08/10/2022
					CT145	Cardiovascular devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	08/10/2022	08/10/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	08/10/2022	08/10/2022
					CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	08/10/2022	08/10/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	08/10/2022	08/10/2022
					CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	08/10/2022	08/10/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	08/10/2022	08/10/2022
					CT233	Surgical	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	08/10/2022	08/10/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	08/10/2022	08/10/2022
					CT2406	Implantable	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	08/10/2022	08/10/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	08/10/2022	08/10/2022
					CT2407	Active implantable	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	08/10/2022	08/10/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	08/10/2022	08/10/2022
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	08/10/2022	08/10/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	08/10/2022	08/10/2022
					CT2727	Active implant devices/ancillaries	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	08/10/2022	08/10/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	08/10/2022	08/10/2022
					CT2750	Implantable cardiac electrical stimulators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	08/10/2022	08/10/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	08/10/2022	08/10/2022
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	08/10/2022	08/10/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	08/10/2022	08/10/2022
					CT2985	Active	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	08/10/2022	08/10/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	08/10/2022	08/10/2022
					CT3001	Implantable metal	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	08/10/2022	08/10/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	08/10/2022	08/10/2022
					CT3002	Implantable synthetic polymer	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	08/10/2022	08/10/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	08/10/2022	08/10/2022
					CT3006	Implantable inorganic material	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	08/10/2022	08/10/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	08/10/2022	08/10/2022
					CT314	Electrically powered	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	08/10/2022	08/10/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	08/10/2022	08/10/2022
					CT321	Long-term surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	08/10/2022	08/10/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	08/10/2022	08/10/2022
					CT449	Generators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	08/10/2022	08/10/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	08/10/2022	08/10/2022
					CT981	Single use	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	08/10/2022	08/10/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	08/10/2022	08/10/2022

					CT983	Surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	08/10/2022	08/10/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	08/10/2022	08/10/2022
11988	2023	All other reportable incidents	35223	Endocardial/interventricular septal pacing lead	CT1684	Cardiac pacemakers and associated devices	E0112	Dizziness	A070101	Failure to Capture		
							E011903	Syncope/Fainting	A070101	Failure to Capture		
							F23	Unexpected Medical Intervention	A070101	Failure to Capture		
11989	2023		47265	Dual-chamber implantable pacemaker, rate-responsive	CT141	Electrodes/Leads and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem		
							F1901	Additional Surgery	A24	No Apparent Device Problem		
					CT145	Cardiovascular devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem		
							F1901	Additional Surgery	A24	No Apparent Device Problem		
					CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem		
							F1901	Additional Surgery	A24	No Apparent Device Problem		
					CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem		
							F1901	Additional Surgery	A24	No Apparent Device Problem		
					CT233	Surgical	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem		
							F1901	Additional Surgery	A24	No Apparent Device Problem		
					CT2406	Implantable	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem		
							F1901	Additional Surgery	A24	No Apparent Device Problem		
					CT2407	Active implantable	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem		
							F1901	Additional Surgery	A24	No Apparent Device Problem		
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem		
							F1901	Additional Surgery	A24	No Apparent Device Problem		
					CT2727	Active implant devices/ancillaries	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem		
							F1901	Additional Surgery	A24	No Apparent Device Problem		
					CT2750	Implantable cardiac electrical stimulators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem		
							F1901	Additional Surgery	A24	No Apparent Device Problem		
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem		
							F1901	Additional Surgery	A24	No Apparent Device Problem		
					CT2985	Active	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem		
							F1901	Additional Surgery	A24	No Apparent Device Problem		
					CT3001	Implantable metal	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem		
							F1901	Additional Surgery	A24	No Apparent Device Problem		
					CT3002	Implantable synthetic polymer	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem		
							F1901	Additional Surgery	A24	No Apparent Device Problem		
					CT3006	Implantable inorganic material	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem		
							F1901	Additional Surgery	A24	No Apparent Device Problem		
					CT314	Electrically powered	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem		
							F1901	Additional Surgery	A24	No Apparent Device Problem		
					CT321	Long-term surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem		
							F1901	Additional Surgery	A24	No Apparent Device Problem		
					CT449	Generators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem		
							F1901	Additional Surgery	A24	No Apparent Device Problem		
					CT981	Single use	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem		
							F1901	Additional Surgery	A24	No Apparent Device Problem		
					CT983	Surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem		
							F1901	Additional Surgery	A24	No Apparent Device Problem		
11990	2023		47265	Dual-chamber implantable pacemaker, rate-responsive	CT141	Electrodes/Leads and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem		
							F1901	Additional Surgery	A24	No Apparent Device Problem		
					CT145	Cardiovascular devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem		
							F1901	Additional Surgery	A24	No Apparent Device Problem		
					CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem		
							F1901	Additional Surgery	A24	No Apparent Device Problem		
					CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem		
							F1901	Additional Surgery	A24	No Apparent Device Problem		
					CT233	Surgical	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem		
							F1901	Additional Surgery	A24	No Apparent Device Problem		
					CT2406	Implantable	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem		
							F1901	Additional Surgery	A24	No Apparent Device Problem		
					CT2407	Active implantable	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem		
							F1901	Additional Surgery	A24	No Apparent Device Problem		
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem		
							F1901	Additional Surgery	A24	No Apparent Device Problem		
					CT2727	Active implant devices/ancillaries	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem		
							F1901	Additional Surgery	A24	No Apparent Device Problem		
					CT2750	Implantable cardiac electrical stimulators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem		
							F1901	Additional Surgery	A24	No Apparent Device Problem		
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem		
							F1901	Additional Surgery	A24	No Apparent Device Problem		
					CT2985	Active	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem		
							F1901	Additional Surgery	A24	No Apparent Device Problem		
					CT3001	Implantable metal	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem		
							F1901	Additional Surgery	A24	No Apparent Device Problem		
					CT3002	Implantable synthetic polymer	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem		
							F1901	Additional Surgery	A24	No Apparent Device Problem		
					CT3006	Implantable inorganic material	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem		
							F1901	Additional Surgery	A24	No Apparent Device Problem		
					CT314	Electrically powered	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem		
							F1901	Additional Surgery	A24	No Apparent Device Problem		
					CT321	Long-term surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem		
							F1901	Additional Surgery	A24	No Apparent Device Problem		
					CT449	Generators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem		
							F1901	Additional Surgery	A24	No Apparent Device Problem		
					CT981	Single use	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem		
							F1901	Additional Surgery	A24	No Apparent Device Problem		
					CT983	Surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem		
							F1901	Additional Surgery	A24	No Apparent Device Problem		
11991	2023		35223	Endocardial/interventricular septal pacing lead	CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A25	No Apparent Adverse Event		
							F26	No Health Consequences or Impact	A25	No Apparent Adverse Event		
11992	2023		47265	Dual-chamber implantable pacemaker, rate-	CT141	Electrodes/Leads and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	09/01/2023	09/01/2023

				responsive			F1901	Additional Surgery	A24	No Apparent Device Problem	09/01/2023	09/01/2023
					CT145	Cardiovascular devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	09/01/2023	09/01/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	09/01/2023	09/01/2023
					CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	09/01/2023	09/01/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	09/01/2023	09/01/2023
					CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	09/01/2023	09/01/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	09/01/2023	09/01/2023
					CT233	Surgical	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	09/01/2023	09/01/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	09/01/2023	09/01/2023
					CT2406	Implantable	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	09/01/2023	09/01/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	09/01/2023	09/01/2023
					CT2407	Active implantable	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	09/01/2023	09/01/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	09/01/2023	09/01/2023
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	09/01/2023	09/01/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	09/01/2023	09/01/2023
					CT2727	Active implant devices/ancillaries	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	09/01/2023	09/01/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	09/01/2023	09/01/2023
					CT2750	Implantable cardiac electrical stimulators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	09/01/2023	09/01/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	09/01/2023	09/01/2023
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	09/01/2023	09/01/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	09/01/2023	09/01/2023
					CT2985	Active	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	09/01/2023	09/01/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	09/01/2023	09/01/2023
					CT3001	Implantable metal	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	09/01/2023	09/01/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	09/01/2023	09/01/2023
					CT3002	Implantable synthetic polymer	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	09/01/2023	09/01/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	09/01/2023	09/01/2023
					CT3006	Implantable inorganic material	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	09/01/2023	09/01/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	09/01/2023	09/01/2023
					CT314	Electrically powered	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	09/01/2023	09/01/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	09/01/2023	09/01/2023
					CT321	Long-term surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	09/01/2023	09/01/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	09/01/2023	09/01/2023
					CT449	Generators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	09/01/2023	09/01/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	09/01/2023	09/01/2023
					CT981	Single use	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	09/01/2023	09/01/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	09/01/2023	09/01/2023
					CT983	Surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	09/01/2023	09/01/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	09/01/2023	09/01/2023
11993	2023		47265	Dual-chamber implantable pacemaker, rate-responsive	CT141	Electrodes/Leads and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	05/01/2023	05/01/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	05/01/2023	05/01/2023
					CT145	Cardiovascular devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	05/01/2023	05/01/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	05/01/2023	05/01/2023
					CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	05/01/2023	05/01/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	05/01/2023	05/01/2023
					CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	05/01/2023	05/01/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	05/01/2023	05/01/2023
					CT233	Surgical	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	05/01/2023	05/01/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	05/01/2023	05/01/2023
					CT2406	Implantable	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	05/01/2023	05/01/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	05/01/2023	05/01/2023
					CT2407	Active implantable	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	05/01/2023	05/01/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	05/01/2023	05/01/2023
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	05/01/2023	05/01/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	05/01/2023	05/01/2023
					CT2727	Active implant devices/ancillaries	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	05/01/2023	05/01/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	05/01/2023	05/01/2023
					CT2750	Implantable cardiac electrical stimulators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	05/01/2023	05/01/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	05/01/2023	05/01/2023
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	05/01/2023	05/01/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	05/01/2023	05/01/2023
					CT2985	Active	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	05/01/2023	05/01/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	05/01/2023	05/01/2023
					CT3001	Implantable metal	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	05/01/2023	05/01/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	05/01/2023	05/01/2023
					CT3002	Implantable synthetic polymer	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	05/01/2023	05/01/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	05/01/2023	05/01/2023
					CT3006	Implantable inorganic material	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	05/01/2023	05/01/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	05/01/2023	05/01/2023
					CT314	Electrically powered	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	05/01/2023	05/01/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	05/01/2023	05/01/2023
					CT321	Long-term surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	05/01/2023	05/01/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	05/01/2023	05/01/2023
					CT449	Generators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	05/01/2023	05/01/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	05/01/2023	05/01/2023
					CT981	Single use	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	05/01/2023	05/01/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	05/01/2023	05/01/2023
					CT983	Surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	05/01/2023	05/01/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	05/01/2023	05/01/2023
11994	2023	All other reportable incidents	47263	Cardiac resynchronization therapy implantable pacemaker	CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A0908	Incorrect, Inadequate or Imprecise Result or Readings	29/12/2022	29/12/2022
							F08	Hospitalization or Prolonged Hospitalization	A0908	Incorrect, Inadequate or Imprecise Result or Readings	29/12/2022	29/12/2022
							F1905	Device Revision or Replacement	A0908	Incorrect, Inadequate or Imprecise Result or Readings	29/12/2022	29/12/2022
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A0908	Incorrect, Inadequate or Imprecise Result or Readings	29/12/2022	29/12/2022
							F08	Hospitalization or Prolonged Hospitalization	A0908	Incorrect, Inadequate or Imprecise Result or Readings	29/12/2022	29/12/2022

							F1905	Device Revision or Replacement	A0908	Incorrect, Inadequate or Imprecise Result or Readings	29/12/2022	29/12/2022
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A0908	Incorrect, Inadequate or Imprecise Result or Readings	29/12/2022	29/12/2022
							F08	Hospitalization or Prolonged Hospitalization	A0908	Incorrect, Inadequate or Imprecise Result or Readings	29/12/2022	29/12/2022
							F1905	Device Revision or Replacement	A0908	Incorrect, Inadequate or Imprecise Result or Readings	29/12/2022	29/12/2022
11995	2023	All other reportable incidents	47265	Dual-chamber implantable pacemaker, rate-responsive	CT141	Electrodes/Leads and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	12/11/2022	12/11/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	12/11/2022	12/11/2022
					CT145	Cardiovascular devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	12/11/2022	12/11/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	12/11/2022	12/11/2022
					CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	12/11/2022	12/11/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	12/11/2022	12/11/2022
					CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	12/11/2022	12/11/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	12/11/2022	12/11/2022
					CT233	Surgical	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	12/11/2022	12/11/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	12/11/2022	12/11/2022
					CT2406	Implantable	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	12/11/2022	12/11/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	12/11/2022	12/11/2022
					CT2407	Active implantable	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	12/11/2022	12/11/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	12/11/2022	12/11/2022
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	12/11/2022	12/11/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	12/11/2022	12/11/2022
					CT2727	Active implant devices/ancillaries	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	12/11/2022	12/11/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	12/11/2022	12/11/2022
					CT2750	Implantable cardiac electrical stimulators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	12/11/2022	12/11/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	12/11/2022	12/11/2022
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	12/11/2022	12/11/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	12/11/2022	12/11/2022
					CT2985	Active	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	12/11/2022	12/11/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	12/11/2022	12/11/2022
					CT3001	Implantable metal	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	12/11/2022	12/11/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	12/11/2022	12/11/2022
					CT3002	Implantable synthetic polymer	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	12/11/2022	12/11/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	12/11/2022	12/11/2022
					CT3006	Implantable inorganic material	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	12/11/2022	12/11/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	12/11/2022	12/11/2022
					CT314	Electrically powered	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	12/11/2022	12/11/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	12/11/2022	12/11/2022
					CT321	Long-term surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	12/11/2022	12/11/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	12/11/2022	12/11/2022
					CT449	Generators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	12/11/2022	12/11/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	12/11/2022	12/11/2022
					CT981	Single use	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	12/11/2022	12/11/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	12/11/2022	12/11/2022
					CT983	Surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	12/11/2022	12/11/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	12/11/2022	12/11/2022
11996	2023	All other reportable incidents	47265	Dual-chamber implantable pacemaker, rate-responsive	CT141	Electrodes/Leads and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem		
							F1901	Additional Surgery	A24	No Apparent Device Problem		
					CT145	Cardiovascular devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem		
							F1901	Additional Surgery	A24	No Apparent Device Problem		
					CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem		
							F1901	Additional Surgery	A24	No Apparent Device Problem		
					CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem		
							F1901	Additional Surgery	A24	No Apparent Device Problem		
					CT233	Surgical	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem		
							F1901	Additional Surgery	A24	No Apparent Device Problem		
					CT2406	Implantable	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem		
							F1901	Additional Surgery	A24	No Apparent Device Problem		
					CT2407	Active implantable	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem		
							F1901	Additional Surgery	A24	No Apparent Device Problem		
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem		
							F1901	Additional Surgery	A24	No Apparent Device Problem		
					CT2727	Active implant devices/ancillaries	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem		
							F1901	Additional Surgery	A24	No Apparent Device Problem		
					CT2750	Implantable cardiac electrical stimulators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem		
							F1901	Additional Surgery	A24	No Apparent Device Problem		
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem		
							F1901	Additional Surgery	A24	No Apparent Device Problem		
					CT2985	Active	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem		
							F1901	Additional Surgery	A24	No Apparent Device Problem		
					CT3001	Implantable metal	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem		
							F1901	Additional Surgery	A24	No Apparent Device Problem		
					CT3002	Implantable synthetic polymer	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem		
							F1901	Additional Surgery	A24	No Apparent Device Problem		
					CT3006	Implantable inorganic material	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem		
							F1901	Additional Surgery	A24	No Apparent Device Problem		
					CT314	Electrically powered	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem		
							F1901	Additional Surgery	A24	No Apparent Device Problem		
					CT321	Long-term surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem		
							F1901	Additional Surgery	A24	No Apparent Device Problem		
					CT449	Generators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem		
							F1901	Additional Surgery	A24	No Apparent Device Problem		
					CT981	Single use	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem		
							F1901	Additional Surgery	A24	No Apparent Device Problem		
					CT983	Surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem		
							F1901	Additional Surgery	A24	No Apparent Device Problem		
11997	2023		47265	Dual-chamber implantable pacemaker, rate-	CT141	Electrodes/Leads and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	07/01/2023	07/01/2023

				responsive			F1901	Additional Surgery	A24	No Apparent Device Problem	07/01/2023	07/01/2023
					CT145	Cardiovascular devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	07/01/2023	07/01/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	07/01/2023	07/01/2023
					CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	07/01/2023	07/01/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	07/01/2023	07/01/2023
					CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	07/01/2023	07/01/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	07/01/2023	07/01/2023
					CT233	Surgical	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	07/01/2023	07/01/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	07/01/2023	07/01/2023
					CT2406	Implantable	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	07/01/2023	07/01/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	07/01/2023	07/01/2023
					CT2407	Active implantable	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	07/01/2023	07/01/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	07/01/2023	07/01/2023
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	07/01/2023	07/01/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	07/01/2023	07/01/2023
					CT2727	Active implant devices/ancillaries	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	07/01/2023	07/01/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	07/01/2023	07/01/2023
					CT2750	Implantable cardiac electrical stimulators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	07/01/2023	07/01/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	07/01/2023	07/01/2023
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	07/01/2023	07/01/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	07/01/2023	07/01/2023
					CT2985	Active	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	07/01/2023	07/01/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	07/01/2023	07/01/2023
					CT3001	Implantable metal	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	07/01/2023	07/01/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	07/01/2023	07/01/2023
					CT3002	Implantable synthetic polymer	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	07/01/2023	07/01/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	07/01/2023	07/01/2023
					CT3006	Implantable inorganic material	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	07/01/2023	07/01/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	07/01/2023	07/01/2023
					CT314	Electrically powered	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	07/01/2023	07/01/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	07/01/2023	07/01/2023
					CT321	Long-term surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	07/01/2023	07/01/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	07/01/2023	07/01/2023
					CT449	Generators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	07/01/2023	07/01/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	07/01/2023	07/01/2023
					CT981	Single use	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	07/01/2023	07/01/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	07/01/2023	07/01/2023
					CT983	Surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	07/01/2023	07/01/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	07/01/2023	07/01/2023
11998	2023	All other reportable incidents	17882	Physiologic-monitoring defibrillation system	CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070803	Failure to Power Up		
							F27	Problem Identified During Non-Clinical Procedure	A070803	Failure to Power Up		
11999	2023	All other reportable incidents	17882	Physiologic-monitoring defibrillation system	CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A0708	Power Problem		
							F27	Problem Identified During Non-Clinical Procedure	A0708	Power Problem		
12000	2023	All other reportable incidents	35852	Single-chamber implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E2104	Electric Shock	A0508	Noise, Audible		
									A070909	Oversensing		
									A071302	Inappropriate/Inadequate Shock/Stimulation		
							F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A0508	Noise, Audible		
									A070909	Oversensing		
									A071302	Inappropriate/Inadequate Shock/Stimulation		
							F15	Recognized Device or Procedural Complication	A0508	Noise, Audible		
									A070909	Oversensing		
									A071302	Inappropriate/Inadequate Shock/Stimulation		
12001	2023	All other reportable incidents	35852	Single-chamber implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E2104	Electric Shock	A0508	Noise, Audible		
									A070909	Oversensing		
									A071302	Inappropriate/Inadequate Shock/Stimulation		
							F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A0508	Noise, Audible		
									A070909	Oversensing		
									A071302	Inappropriate/Inadequate Shock/Stimulation		
							F15	Recognized Device or Procedural Complication	A0508	Noise, Audible		
									A070909	Oversensing		
									A071302	Inappropriate/Inadequate Shock/Stimulation		
12002	2023		47265	Dual-chamber implantable pacemaker, rate-responsive	CT141	Electrodes/Leads and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	07/12/2022	07/12/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	07/12/2022	07/12/2022
					CT145	Cardiovascular devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	07/12/2022	07/12/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	07/12/2022	07/12/2022
					CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	07/12/2022	07/12/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	07/12/2022	07/12/2022
					CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	07/12/2022	07/12/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	07/12/2022	07/12/2022
					CT233	Surgical	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	07/12/2022	07/12/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	07/12/2022	07/12/2022
					CT2406	Implantable	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	07/12/2022	07/12/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	07/12/2022	07/12/2022
					CT2407	Active implantable	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	07/12/2022	07/12/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	07/12/2022	07/12/2022
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	07/12/2022	07/12/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	07/12/2022	07/12/2022
					CT2727	Active implant devices/ancillaries	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	07/12/2022	07/12/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	07/12/2022	07/12/2022
					CT2750	Implantable cardiac electrical stimulators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	07/12/2022	07/12/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	07/12/2022	07/12/2022
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	07/12/2022	07/12/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	07/12/2022	07/12/2022
					CT2985	Active	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	07/12/2022	07/12/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	07/12/2022	07/12/2022
					CT3001	Implantable metal	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	07/12/2022	07/12/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	07/12/2022	07/12/2022
					CT3002	Implantable synthetic polymer	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	07/12/2022	07/12/2022

						F1901	Additional Surgery	A24	No Apparent Device Problem	07/12/2022	07/12/2022
				CT3006	Implantable inorganic material	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	07/12/2022	07/12/2022
						F1901	Additional Surgery	A24	No Apparent Device Problem	07/12/2022	07/12/2022
				CT314	Electrically powered	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	07/12/2022	07/12/2022
						F1901	Additional Surgery	A24	No Apparent Device Problem	07/12/2022	07/12/2022
				CT321	Long-term surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	07/12/2022	07/12/2022
						F1901	Additional Surgery	A24	No Apparent Device Problem	07/12/2022	07/12/2022
				CT449	Generators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	07/12/2022	07/12/2022
						F1901	Additional Surgery	A24	No Apparent Device Problem	07/12/2022	07/12/2022
				CT981	Single use	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	07/12/2022	07/12/2022
						F1901	Additional Surgery	A24	No Apparent Device Problem	07/12/2022	07/12/2022
				CT983	Surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	07/12/2022	07/12/2022
						F1901	Additional Surgery	A24	No Apparent Device Problem	07/12/2022	07/12/2022
12003	2023	All other reportable incidents	47270	Cardiac resynchronization therapy implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem	04/11/2022	04/11/2022
									A070504	Premature Discharge of Battery	04/11/2022	04/11/2022
							F08	Hospitalization or Prolonged Hospitalization	A0705	Battery Problem	04/11/2022	04/11/2022
									A070504	Premature Discharge of Battery	04/11/2022	04/11/2022
							F1905	Device Revision or Replacement	A0705	Battery Problem	04/11/2022	04/11/2022
									A070504	Premature Discharge of Battery	04/11/2022	04/11/2022
12004	2023	All other reportable incidents	47265	Dual-chamber implantable pacemaker, rate-responsive	CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	12/11/2021	12/11/2021
							F1901	Additional Surgery	A24	No Apparent Device Problem	12/11/2021	12/11/2021
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	12/11/2021	12/11/2021
							F1901	Additional Surgery	A24	No Apparent Device Problem	12/11/2021	12/11/2021
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	12/11/2021	12/11/2021
							F1901	Additional Surgery	A24	No Apparent Device Problem	12/11/2021	12/11/2021
12005	2023		47265	Dual-chamber implantable pacemaker, rate-responsive	CT141	Electrodes/Leads and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	18/10/2022	18/10/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	18/10/2022	18/10/2022
					CT145	Cardiovascular devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	18/10/2022	18/10/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	18/10/2022	18/10/2022
					CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	18/10/2022	18/10/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	18/10/2022	18/10/2022
					CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	18/10/2022	18/10/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	18/10/2022	18/10/2022
					CT233	Surgical	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	18/10/2022	18/10/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	18/10/2022	18/10/2022
					CT2406	Implantable	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	18/10/2022	18/10/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	18/10/2022	18/10/2022
					CT2407	Active implantable	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	18/10/2022	18/10/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	18/10/2022	18/10/2022
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	18/10/2022	18/10/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	18/10/2022	18/10/2022
					CT2727	Active implant devices/ancillaries	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	18/10/2022	18/10/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	18/10/2022	18/10/2022
					CT2750	Implantable cardiac electrical stimulators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	18/10/2022	18/10/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	18/10/2022	18/10/2022
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	18/10/2022	18/10/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	18/10/2022	18/10/2022
					CT2985	Active	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	18/10/2022	18/10/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	18/10/2022	18/10/2022
					CT3001	Implantable metal	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	18/10/2022	18/10/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	18/10/2022	18/10/2022
					CT3002	Implantable synthetic polymer	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	18/10/2022	18/10/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	18/10/2022	18/10/2022
					CT3006	Implantable inorganic material	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	18/10/2022	18/10/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	18/10/2022	18/10/2022
					CT314	Electrically powered	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	18/10/2022	18/10/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	18/10/2022	18/10/2022
					CT321	Long-term surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	18/10/2022	18/10/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	18/10/2022	18/10/2022
					CT449	Generators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	18/10/2022	18/10/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	18/10/2022	18/10/2022
					CT981	Single use	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	18/10/2022	18/10/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	18/10/2022	18/10/2022
					CT983	Surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	18/10/2022	18/10/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	18/10/2022	18/10/2022
12006	2023		47265	Dual-chamber implantable pacemaker, rate-responsive	CT141	Electrodes/Leads and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	05/01/2023	05/01/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	05/01/2023	05/01/2023
					CT145	Cardiovascular devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	05/01/2023	05/01/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	05/01/2023	05/01/2023
					CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	05/01/2023	05/01/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	05/01/2023	05/01/2023
					CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	05/01/2023	05/01/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	05/01/2023	05/01/2023
					CT233	Surgical	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	05/01/2023	05/01/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	05/01/2023	05/01/2023
					CT2406	Implantable	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	05/01/2023	05/01/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	05/01/2023	05/01/2023
					CT2407	Active implantable	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	05/01/2023	05/01/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	05/01/2023	05/01/2023
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	05/01/2023	05/01/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	05/01/2023	05/01/2023
					CT2727	Active implant devices/ancillaries	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	05/01/2023	05/01/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	05/01/2023	05/01/2023
					CT2750	Implantable cardiac electrical stimulators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	05/01/2023	05/01/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	05/01/2023	05/01/2023
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	05/01/2023	05/01/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	05/01/2023	05/01/2023
					CT2985	Active	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	05/01/2023	05/01/2023

							F1901	Additional Surgery	A24	No Apparent Device Problem	05/01/2023	05/01/2023
					CT3001	Implantable metal	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	05/01/2023	05/01/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	05/01/2023	05/01/2023
					CT3002	Implantable synthetic polymer	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	05/01/2023	05/01/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	05/01/2023	05/01/2023
					CT3006	Implantable inorganic material	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	05/01/2023	05/01/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	05/01/2023	05/01/2023
					CT314	Electrically powered	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	05/01/2023	05/01/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	05/01/2023	05/01/2023
					CT321	Long-term surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	05/01/2023	05/01/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	05/01/2023	05/01/2023
					CT449	Generators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	05/01/2023	05/01/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	05/01/2023	05/01/2023
					CT981	Single use	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	05/01/2023	05/01/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	05/01/2023	05/01/2023
					CT983	Surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	05/01/2023	05/01/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	05/01/2023	05/01/2023
12007	2023	All other reportable incidents	47265	Dual-chamber implantable pacemaker, rate-responsive	CT141	Electrodes/Leads and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	20/12/2022	20/12/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	20/12/2022	20/12/2022
					CT145	Cardiovascular devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	20/12/2022	20/12/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	20/12/2022	20/12/2022
					CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	20/12/2022	20/12/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	20/12/2022	20/12/2022
					CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	20/12/2022	20/12/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	20/12/2022	20/12/2022
					CT233	Surgical	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	20/12/2022	20/12/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	20/12/2022	20/12/2022
					CT2406	Implantable	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	20/12/2022	20/12/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	20/12/2022	20/12/2022
					CT2407	Active implantable	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	20/12/2022	20/12/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	20/12/2022	20/12/2022
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	20/12/2022	20/12/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	20/12/2022	20/12/2022
					CT2727	Active implant devices/ancillaries	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	20/12/2022	20/12/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	20/12/2022	20/12/2022
					CT2750	Implantable cardiac electrical stimulators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	20/12/2022	20/12/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	20/12/2022	20/12/2022
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	20/12/2022	20/12/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	20/12/2022	20/12/2022
					CT2985	Active	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	20/12/2022	20/12/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	20/12/2022	20/12/2022
					CT3001	Implantable metal	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	20/12/2022	20/12/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	20/12/2022	20/12/2022
					CT3002	Implantable synthetic polymer	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	20/12/2022	20/12/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	20/12/2022	20/12/2022
					CT3006	Implantable inorganic material	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	20/12/2022	20/12/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	20/12/2022	20/12/2022
					CT314	Electrically powered	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	20/12/2022	20/12/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	20/12/2022	20/12/2022
					CT321	Long-term surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	20/12/2022	20/12/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	20/12/2022	20/12/2022
					CT449	Generators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	20/12/2022	20/12/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	20/12/2022	20/12/2022
					CT981	Single use	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	20/12/2022	20/12/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	20/12/2022	20/12/2022
					CT983	Surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	20/12/2022	20/12/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	20/12/2022	20/12/2022
12008	2023	All other reportable incidents	47265	Dual-chamber implantable pacemaker, rate-responsive	CT141	Electrodes/Leads and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	12/11/2022	12/11/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	12/11/2022	12/11/2022
					CT145	Cardiovascular devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	12/11/2022	12/11/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	12/11/2022	12/11/2022
					CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	12/11/2022	12/11/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	12/11/2022	12/11/2022
					CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	12/11/2022	12/11/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	12/11/2022	12/11/2022
					CT233	Surgical	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	12/11/2022	12/11/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	12/11/2022	12/11/2022
					CT2406	Implantable	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	12/11/2022	12/11/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	12/11/2022	12/11/2022
					CT2407	Active implantable	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	12/11/2022	12/11/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	12/11/2022	12/11/2022
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	12/11/2022	12/11/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	12/11/2022	12/11/2022
					CT2727	Active implant devices/ancillaries	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	12/11/2022	12/11/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	12/11/2022	12/11/2022
					CT2750	Implantable cardiac electrical stimulators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	12/11/2022	12/11/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	12/11/2022	12/11/2022
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	12/11/2022	12/11/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	12/11/2022	12/11/2022
					CT2985	Active	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	12/11/2022	12/11/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	12/11/2022	12/11/2022
					CT3001	Implantable metal	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	12/11/2022	12/11/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	12/11/2022	12/11/2022
					CT3002	Implantable synthetic polymer	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	12/11/2022	12/11/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	12/11/2022	12/11/2022
					CT3006	Implantable inorganic material	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	12/11/2022	12/11/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	12/11/2022	12/11/2022
					CT314	Electrically powered	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	12/11/2022	12/11/2022

							F1901	Additional Surgery	A24	No Apparent Device Problem	12/11/2022	12/11/2022
					CT321	Long-term surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	12/11/2022	12/11/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	12/11/2022	12/11/2022
					CT449	Generators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	12/11/2022	12/11/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	12/11/2022	12/11/2022
					CT981	Single use	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	12/11/2022	12/11/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	12/11/2022	12/11/2022
					CT983	Surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	12/11/2022	12/11/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	12/11/2022	12/11/2022
12009	2023		47265	Dual-chamber implantable pacemaker, rate-responsive	CT141	Electrodes/Leads and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	22/12/2022	22/12/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	22/12/2022	22/12/2022
					CT145	Cardiovascular devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	22/12/2022	22/12/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	22/12/2022	22/12/2022
					CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	22/12/2022	22/12/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	22/12/2022	22/12/2022
					CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	22/12/2022	22/12/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	22/12/2022	22/12/2022
					CT233	Surgical	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	22/12/2022	22/12/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	22/12/2022	22/12/2022
					CT2406	Implantable	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	22/12/2022	22/12/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	22/12/2022	22/12/2022
					CT2407	Active implantable	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	22/12/2022	22/12/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	22/12/2022	22/12/2022
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	22/12/2022	22/12/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	22/12/2022	22/12/2022
					CT2727	Active implant devices/ancillaries	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	22/12/2022	22/12/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	22/12/2022	22/12/2022
					CT2750	Implantable cardiac electrical stimulators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	22/12/2022	22/12/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	22/12/2022	22/12/2022
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	22/12/2022	22/12/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	22/12/2022	22/12/2022
					CT2985	Active	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	22/12/2022	22/12/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	22/12/2022	22/12/2022
					CT3001	Implantable metal	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	22/12/2022	22/12/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	22/12/2022	22/12/2022
					CT3002	Implantable synthetic polymer	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	22/12/2022	22/12/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	22/12/2022	22/12/2022
					CT3006	Implantable inorganic material	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	22/12/2022	22/12/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	22/12/2022	22/12/2022
					CT314	Electrically powered	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	22/12/2022	22/12/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	22/12/2022	22/12/2022
					CT321	Long-term surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	22/12/2022	22/12/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	22/12/2022	22/12/2022
					CT449	Generators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	22/12/2022	22/12/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	22/12/2022	22/12/2022
					CT981	Single use	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	22/12/2022	22/12/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	22/12/2022	22/12/2022
					CT983	Surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	22/12/2022	22/12/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	22/12/2022	22/12/2022
12010	2023		47265	Dual-chamber implantable pacemaker, rate-responsive	CT141	Electrodes/Leads and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	18/11/2022	18/11/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	18/11/2022	18/11/2022
					CT145	Cardiovascular devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	18/11/2022	18/11/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	18/11/2022	18/11/2022
					CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	18/11/2022	18/11/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	18/11/2022	18/11/2022
					CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	18/11/2022	18/11/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	18/11/2022	18/11/2022
					CT233	Surgical	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	18/11/2022	18/11/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	18/11/2022	18/11/2022
					CT2406	Implantable	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	18/11/2022	18/11/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	18/11/2022	18/11/2022
					CT2407	Active implantable	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	18/11/2022	18/11/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	18/11/2022	18/11/2022
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	18/11/2022	18/11/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	18/11/2022	18/11/2022
					CT2727	Active implant devices/ancillaries	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	18/11/2022	18/11/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	18/11/2022	18/11/2022
					CT2750	Implantable cardiac electrical stimulators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	18/11/2022	18/11/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	18/11/2022	18/11/2022
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	18/11/2022	18/11/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	18/11/2022	18/11/2022
					CT2985	Active	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	18/11/2022	18/11/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	18/11/2022	18/11/2022
					CT3001	Implantable metal	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	18/11/2022	18/11/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	18/11/2022	18/11/2022
					CT3002	Implantable synthetic polymer	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	18/11/2022	18/11/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	18/11/2022	18/11/2022
					CT3006	Implantable inorganic material	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	18/11/2022	18/11/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	18/11/2022	18/11/2022
					CT314	Electrically powered	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	18/11/2022	18/11/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	18/11/2022	18/11/2022
					CT321	Long-term surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	18/11/2022	18/11/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	18/11/2022	18/11/2022
					CT449	Generators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	18/11/2022	18/11/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	18/11/2022	18/11/2022
					CT981	Single use	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	18/11/2022	18/11/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	18/11/2022	18/11/2022
					CT983	Surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	18/11/2022	18/11/2022

							F1901	Additional Surgery	A24	No Apparent Device Problem	18/11/2022	18/11/2022
12011	2023		47265	Dual-chamber implantable pacemaker, rate-responsive	CT141	Electrodes/Leads and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	19/12/2022	19/12/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	19/12/2022	19/12/2022
					CT145	Cardiovascular devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	19/12/2022	19/12/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	19/12/2022	19/12/2022
					CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	19/12/2022	19/12/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	19/12/2022	19/12/2022
					CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	19/12/2022	19/12/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	19/12/2022	19/12/2022
					CT233	Surgical	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	19/12/2022	19/12/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	19/12/2022	19/12/2022
					CT2406	Implantable	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	19/12/2022	19/12/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	19/12/2022	19/12/2022
					CT2407	Active implantable	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	19/12/2022	19/12/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	19/12/2022	19/12/2022
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	19/12/2022	19/12/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	19/12/2022	19/12/2022
					CT2727	Active implant devices/ancillaries	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	19/12/2022	19/12/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	19/12/2022	19/12/2022
					CT2750	Implantable cardiac electrical stimulators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	19/12/2022	19/12/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	19/12/2022	19/12/2022
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	19/12/2022	19/12/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	19/12/2022	19/12/2022
					CT2985	Active	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	19/12/2022	19/12/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	19/12/2022	19/12/2022
					CT3001	Implantable metal	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	19/12/2022	19/12/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	19/12/2022	19/12/2022
					CT3002	Implantable synthetic polymer	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	19/12/2022	19/12/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	19/12/2022	19/12/2022
					CT3006	Implantable inorganic material	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	19/12/2022	19/12/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	19/12/2022	19/12/2022
					CT314	Electrically powered	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	19/12/2022	19/12/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	19/12/2022	19/12/2022
					CT321	Long-term surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	19/12/2022	19/12/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	19/12/2022	19/12/2022
					CT449	Generators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	19/12/2022	19/12/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	19/12/2022	19/12/2022
					CT981	Single use	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	19/12/2022	19/12/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	19/12/2022	19/12/2022
					CT983	Surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	19/12/2022	19/12/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	19/12/2022	19/12/2022
12012	2023	All other reportable incidents	47270	Cardiac resynchronization therapy implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
									A071102	Failure to Interrogate		
							F23	Unexpected Medical Intervention	A070504	Premature Discharge of Battery		
									A071102	Failure to Interrogate		
12013	2023	All other reportable incidents	47265	Dual-chamber implantable pacemaker, rate-responsive	CT141	Electrodes/Leads and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	21/10/2022	21/10/2022
							F1901	Additional Surgery	A070504	Premature Discharge of Battery	21/10/2022	21/10/2022
					CT145	Cardiovascular devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	21/10/2022	21/10/2022
							F1901	Additional Surgery	A070504	Premature Discharge of Battery	21/10/2022	21/10/2022
					CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	21/10/2022	21/10/2022
							F1901	Additional Surgery	A070504	Premature Discharge of Battery	21/10/2022	21/10/2022
					CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	21/10/2022	21/10/2022
							F1901	Additional Surgery	A070504	Premature Discharge of Battery	21/10/2022	21/10/2022
					CT233	Surgical	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	21/10/2022	21/10/2022
							F1901	Additional Surgery	A070504	Premature Discharge of Battery	21/10/2022	21/10/2022
					CT2406	Implantable	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	21/10/2022	21/10/2022
							F1901	Additional Surgery	A070504	Premature Discharge of Battery	21/10/2022	21/10/2022
					CT2407	Active implantable	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	21/10/2022	21/10/2022
							F1901	Additional Surgery	A070504	Premature Discharge of Battery	21/10/2022	21/10/2022
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	21/10/2022	21/10/2022
							F1901	Additional Surgery	A070504	Premature Discharge of Battery	21/10/2022	21/10/2022
					CT2727	Active implant devices/ancillaries	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	21/10/2022	21/10/2022
							F1901	Additional Surgery	A070504	Premature Discharge of Battery	21/10/2022	21/10/2022
					CT2750	Implantable cardiac electrical stimulators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	21/10/2022	21/10/2022
							F1901	Additional Surgery	A070504	Premature Discharge of Battery	21/10/2022	21/10/2022
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	21/10/2022	21/10/2022
							F1901	Additional Surgery	A070504	Premature Discharge of Battery	21/10/2022	21/10/2022
					CT2985	Active	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	21/10/2022	21/10/2022
							F1901	Additional Surgery	A070504	Premature Discharge of Battery	21/10/2022	21/10/2022
					CT3001	Implantable metal	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	21/10/2022	21/10/2022
							F1901	Additional Surgery	A070504	Premature Discharge of Battery	21/10/2022	21/10/2022
					CT3002	Implantable synthetic polymer	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	21/10/2022	21/10/2022
							F1901	Additional Surgery	A070504	Premature Discharge of Battery	21/10/2022	21/10/2022
					CT3006	Implantable inorganic material	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	21/10/2022	21/10/2022
							F1901	Additional Surgery	A070504	Premature Discharge of Battery	21/10/2022	21/10/2022
					CT314	Electrically powered	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	21/10/2022	21/10/2022
							F1901	Additional Surgery	A070504	Premature Discharge of Battery	21/10/2022	21/10/2022
					CT321	Long-term surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	21/10/2022	21/10/2022
							F1901	Additional Surgery	A070504	Premature Discharge of Battery	21/10/2022	21/10/2022
					CT449	Generators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	21/10/2022	21/10/2022
							F1901	Additional Surgery	A070504	Premature Discharge of Battery	21/10/2022	21/10/2022
					CT981	Single use	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	21/10/2022	21/10/2022
							F1901	Additional Surgery	A070504	Premature Discharge of Battery	21/10/2022	21/10/2022
					CT983	Surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	21/10/2022	21/10/2022
							F1901	Additional Surgery	A070504	Premature Discharge of Battery	21/10/2022	21/10/2022
12014	2023		47265	Dual-chamber implantable pacemaker, rate-responsive	CT141	Electrodes/Leads and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	05/01/2023	05/01/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	05/01/2023	05/01/2023
					CT145	Cardiovascular devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	05/01/2023	05/01/2023

							F1901	Additional Surgery	A24	No Apparent Device Problem	05/01/2023	05/01/2023
					CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	05/01/2023	05/01/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	05/01/2023	05/01/2023
					CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	05/01/2023	05/01/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	05/01/2023	05/01/2023
					CT233	Surgical	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	05/01/2023	05/01/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	05/01/2023	05/01/2023
					CT2406	Implantable	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	05/01/2023	05/01/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	05/01/2023	05/01/2023
					CT2407	Active implantable	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	05/01/2023	05/01/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	05/01/2023	05/01/2023
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	05/01/2023	05/01/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	05/01/2023	05/01/2023
					CT2727	Active implant devices/ancillaries	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	05/01/2023	05/01/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	05/01/2023	05/01/2023
					CT2750	Implantable cardiac electrical stimulators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	05/01/2023	05/01/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	05/01/2023	05/01/2023
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	05/01/2023	05/01/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	05/01/2023	05/01/2023
					CT2985	Active	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	05/01/2023	05/01/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	05/01/2023	05/01/2023
					CT3001	Implantable metal	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	05/01/2023	05/01/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	05/01/2023	05/01/2023
					CT3002	Implantable synthetic polymer	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	05/01/2023	05/01/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	05/01/2023	05/01/2023
					CT3006	Implantable inorganic material	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	05/01/2023	05/01/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	05/01/2023	05/01/2023
					CT314	Electrically powered	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	05/01/2023	05/01/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	05/01/2023	05/01/2023
					CT321	Long-term surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	05/01/2023	05/01/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	05/01/2023	05/01/2023
					CT449	Generators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	05/01/2023	05/01/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	05/01/2023	05/01/2023
					CT981	Single use	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	05/01/2023	05/01/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	05/01/2023	05/01/2023
					CT983	Surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	05/01/2023	05/01/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	05/01/2023	05/01/2023
12015	2023		47265	Dual-chamber implantable pacemaker, rate-responsive	CT141	Electrodes/Leads and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	05/01/2023	05/01/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	05/01/2023	05/01/2023
					CT145	Cardiovascular devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	05/01/2023	05/01/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	05/01/2023	05/01/2023
					CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	05/01/2023	05/01/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	05/01/2023	05/01/2023
					CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	05/01/2023	05/01/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	05/01/2023	05/01/2023
					CT233	Surgical	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	05/01/2023	05/01/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	05/01/2023	05/01/2023
					CT2406	Implantable	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	05/01/2023	05/01/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	05/01/2023	05/01/2023
					CT2407	Active implantable	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	05/01/2023	05/01/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	05/01/2023	05/01/2023
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	05/01/2023	05/01/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	05/01/2023	05/01/2023
					CT2727	Active implant devices/ancillaries	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	05/01/2023	05/01/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	05/01/2023	05/01/2023
					CT2750	Implantable cardiac electrical stimulators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	05/01/2023	05/01/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	05/01/2023	05/01/2023
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	05/01/2023	05/01/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	05/01/2023	05/01/2023
					CT2985	Active	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	05/01/2023	05/01/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	05/01/2023	05/01/2023
					CT3001	Implantable metal	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	05/01/2023	05/01/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	05/01/2023	05/01/2023
					CT3002	Implantable synthetic polymer	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	05/01/2023	05/01/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	05/01/2023	05/01/2023
					CT3006	Implantable inorganic material	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	05/01/2023	05/01/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	05/01/2023	05/01/2023
					CT314	Electrically powered	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	05/01/2023	05/01/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	05/01/2023	05/01/2023
					CT321	Long-term surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	05/01/2023	05/01/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	05/01/2023	05/01/2023
					CT449	Generators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	05/01/2023	05/01/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	05/01/2023	05/01/2023
					CT981	Single use	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	05/01/2023	05/01/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	05/01/2023	05/01/2023
					CT983	Surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	05/01/2023	05/01/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	05/01/2023	05/01/2023
12016	2023		47265	Dual-chamber implantable pacemaker, rate-responsive	CT141	Electrodes/Leads and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	12/12/2022	12/12/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	12/12/2022	12/12/2022
					CT145	Cardiovascular devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	12/12/2022	12/12/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	12/12/2022	12/12/2022
					CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	12/12/2022	12/12/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	12/12/2022	12/12/2022
					CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	12/12/2022	12/12/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	12/12/2022	12/12/2022
					CT233	Surgical	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	12/12/2022	12/12/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	12/12/2022	12/12/2022
					CT2406	Implantable	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	12/12/2022	12/12/2022

							F1901	Additional Surgery	A24	No Apparent Device Problem	12/12/2022	12/12/2022
					CT2407	Active implantable	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	12/12/2022	12/12/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	12/12/2022	12/12/2022
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	12/12/2022	12/12/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	12/12/2022	12/12/2022
					CT2727	Active implant devices/ancillaries	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	12/12/2022	12/12/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	12/12/2022	12/12/2022
					CT2750	Implantable cardiac electrical stimulators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	12/12/2022	12/12/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	12/12/2022	12/12/2022
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	12/12/2022	12/12/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	12/12/2022	12/12/2022
					CT2985	Active	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	12/12/2022	12/12/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	12/12/2022	12/12/2022
					CT3001	Implantable metal	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	12/12/2022	12/12/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	12/12/2022	12/12/2022
					CT3002	Implantable synthetic polymer	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	12/12/2022	12/12/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	12/12/2022	12/12/2022
					CT3006	Implantable inorganic material	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	12/12/2022	12/12/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	12/12/2022	12/12/2022
					CT314	Electrically powered	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	12/12/2022	12/12/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	12/12/2022	12/12/2022
					CT321	Long-term surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	12/12/2022	12/12/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	12/12/2022	12/12/2022
					CT449	Generators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	12/12/2022	12/12/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	12/12/2022	12/12/2022
					CT981	Single use	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	12/12/2022	12/12/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	12/12/2022	12/12/2022
					CT983	Surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	12/12/2022	12/12/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	12/12/2022	12/12/2022
12017	2023		47265	Dual-chamber implantable pacemaker, rate-responsive	CT141	Electrodes/Leads and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	22/12/2022	22/12/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	22/12/2022	22/12/2022
					CT145	Cardiovascular devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	22/12/2022	22/12/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	22/12/2022	22/12/2022
					CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	22/12/2022	22/12/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	22/12/2022	22/12/2022
					CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	22/12/2022	22/12/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	22/12/2022	22/12/2022
					CT233	Surgical	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	22/12/2022	22/12/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	22/12/2022	22/12/2022
					CT2406	Implantable	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	22/12/2022	22/12/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	22/12/2022	22/12/2022
					CT2407	Active implantable	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	22/12/2022	22/12/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	22/12/2022	22/12/2022
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	22/12/2022	22/12/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	22/12/2022	22/12/2022
					CT2727	Active implant devices/ancillaries	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	22/12/2022	22/12/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	22/12/2022	22/12/2022
					CT2750	Implantable cardiac electrical stimulators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	22/12/2022	22/12/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	22/12/2022	22/12/2022
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	22/12/2022	22/12/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	22/12/2022	22/12/2022
					CT2985	Active	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	22/12/2022	22/12/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	22/12/2022	22/12/2022
					CT3001	Implantable metal	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	22/12/2022	22/12/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	22/12/2022	22/12/2022
					CT3002	Implantable synthetic polymer	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	22/12/2022	22/12/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	22/12/2022	22/12/2022
					CT3006	Implantable inorganic material	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	22/12/2022	22/12/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	22/12/2022	22/12/2022
					CT314	Electrically powered	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	22/12/2022	22/12/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	22/12/2022	22/12/2022
					CT321	Long-term surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	22/12/2022	22/12/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	22/12/2022	22/12/2022
					CT449	Generators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	22/12/2022	22/12/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	22/12/2022	22/12/2022
					CT981	Single use	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	22/12/2022	22/12/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	22/12/2022	22/12/2022
					CT983	Surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	22/12/2022	22/12/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	22/12/2022	22/12/2022
12018	2023		47265	Dual-chamber implantable pacemaker, rate-responsive	CT141	Electrodes/Leads and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	19/12/2022	19/12/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	19/12/2022	19/12/2022
					CT145	Cardiovascular devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	19/12/2022	19/12/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	19/12/2022	19/12/2022
					CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	19/12/2022	19/12/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	19/12/2022	19/12/2022
					CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	19/12/2022	19/12/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	19/12/2022	19/12/2022
					CT233	Surgical	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	19/12/2022	19/12/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	19/12/2022	19/12/2022
					CT2406	Implantable	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	19/12/2022	19/12/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	19/12/2022	19/12/2022
					CT2407	Active implantable	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	19/12/2022	19/12/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	19/12/2022	19/12/2022
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	19/12/2022	19/12/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	19/12/2022	19/12/2022
					CT2727	Active implant devices/ancillaries	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	19/12/2022	19/12/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	19/12/2022	19/12/2022
					CT2750	Implantable cardiac electrical stimulators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	19/12/2022	19/12/2022

							F1901	Additional Surgery	A24	No Apparent Device Problem	19/12/2022	19/12/2022
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	19/12/2022	19/12/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	19/12/2022	19/12/2022
					CT2985	Active	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	19/12/2022	19/12/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	19/12/2022	19/12/2022
					CT3001	Implantable metal	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	19/12/2022	19/12/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	19/12/2022	19/12/2022
					CT3002	Implantable synthetic polymer	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	19/12/2022	19/12/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	19/12/2022	19/12/2022
					CT3006	Implantable inorganic material	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	19/12/2022	19/12/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	19/12/2022	19/12/2022
					CT314	Electrically powered	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	19/12/2022	19/12/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	19/12/2022	19/12/2022
					CT321	Long-term surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	19/12/2022	19/12/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	19/12/2022	19/12/2022
					CT449	Generators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	19/12/2022	19/12/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	19/12/2022	19/12/2022
					CT981	Single use	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	19/12/2022	19/12/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	19/12/2022	19/12/2022
					CT983	Surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	19/12/2022	19/12/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	19/12/2022	19/12/2022
12019	2023		47265	Dual-chamber implantable pacemaker, rate-responsive	CT141	Electrodes/Leads and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	05/01/2023	05/01/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	05/01/2023	05/01/2023
					CT145	Cardiovascular devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	05/01/2023	05/01/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	05/01/2023	05/01/2023
					CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	05/01/2023	05/01/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	05/01/2023	05/01/2023
					CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	05/01/2023	05/01/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	05/01/2023	05/01/2023
					CT233	Surgical	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	05/01/2023	05/01/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	05/01/2023	05/01/2023
					CT2406	Implantable	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	05/01/2023	05/01/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	05/01/2023	05/01/2023
					CT2407	Active implantable	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	05/01/2023	05/01/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	05/01/2023	05/01/2023
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	05/01/2023	05/01/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	05/01/2023	05/01/2023
					CT2727	Active implant devices/ancillaries	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	05/01/2023	05/01/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	05/01/2023	05/01/2023
					CT2750	Implantable cardiac electrical stimulators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	05/01/2023	05/01/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	05/01/2023	05/01/2023
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	05/01/2023	05/01/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	05/01/2023	05/01/2023
					CT2985	Active	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	05/01/2023	05/01/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	05/01/2023	05/01/2023
					CT3001	Implantable metal	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	05/01/2023	05/01/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	05/01/2023	05/01/2023
					CT3002	Implantable synthetic polymer	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	05/01/2023	05/01/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	05/01/2023	05/01/2023
					CT3006	Implantable inorganic material	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	05/01/2023	05/01/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	05/01/2023	05/01/2023
					CT314	Electrically powered	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	05/01/2023	05/01/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	05/01/2023	05/01/2023
					CT321	Long-term surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	05/01/2023	05/01/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	05/01/2023	05/01/2023
					CT449	Generators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	05/01/2023	05/01/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	05/01/2023	05/01/2023
					CT981	Single use	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	05/01/2023	05/01/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	05/01/2023	05/01/2023
					CT983	Surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	05/01/2023	05/01/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	05/01/2023	05/01/2023
12020	2023		47265	Dual-chamber implantable pacemaker, rate-responsive	CT141	Electrodes/Leads and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	30/12/2022	30/12/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	30/12/2022	30/12/2022
					CT145	Cardiovascular devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	30/12/2022	30/12/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	30/12/2022	30/12/2022
					CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	30/12/2022	30/12/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	30/12/2022	30/12/2022
					CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	30/12/2022	30/12/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	30/12/2022	30/12/2022
					CT233	Surgical	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	30/12/2022	30/12/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	30/12/2022	30/12/2022
					CT2406	Implantable	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	30/12/2022	30/12/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	30/12/2022	30/12/2022
					CT2407	Active implantable	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	30/12/2022	30/12/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	30/12/2022	30/12/2022
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	30/12/2022	30/12/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	30/12/2022	30/12/2022
					CT2727	Active implant devices/ancillaries	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	30/12/2022	30/12/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	30/12/2022	30/12/2022
					CT2750	Implantable cardiac electrical stimulators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	30/12/2022	30/12/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	30/12/2022	30/12/2022
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	30/12/2022	30/12/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	30/12/2022	30/12/2022
					CT2985	Active	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	30/12/2022	30/12/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	30/12/2022	30/12/2022
					CT3001	Implantable metal	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	30/12/2022	30/12/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	30/12/2022	30/12/2022
					CT3002	Implantable synthetic polymer	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	30/12/2022	30/12/2022

						F1901	Additional Surgery	A24	No Apparent Device Problem	30/12/2022	30/12/2022
				CT3006	Implantable inorganic material	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	30/12/2022	30/12/2022
						F1901	Additional Surgery	A24	No Apparent Device Problem	30/12/2022	30/12/2022
				CT314	Electrically powered	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	30/12/2022	30/12/2022
						F1901	Additional Surgery	A24	No Apparent Device Problem	30/12/2022	30/12/2022
				CT321	Long-term surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	30/12/2022	30/12/2022
						F1901	Additional Surgery	A24	No Apparent Device Problem	30/12/2022	30/12/2022
				CT449	Generators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	30/12/2022	30/12/2022
						F1901	Additional Surgery	A24	No Apparent Device Problem	30/12/2022	30/12/2022
				CT981	Single use	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	30/12/2022	30/12/2022
						F1901	Additional Surgery	A24	No Apparent Device Problem	30/12/2022	30/12/2022
				CT983	Surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	30/12/2022	30/12/2022
						F1901	Additional Surgery	A24	No Apparent Device Problem	30/12/2022	30/12/2022
12021	2023	All other reportable incidents	47265	Dual-chamber implantable pacemaker, rate-responsive	CT141	Electrodes/Leads and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	12/11/2022	12/11/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	12/11/2022	12/11/2022
					CT145	Cardiovascular devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	12/11/2022	12/11/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	12/11/2022	12/11/2022
					CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	12/11/2022	12/11/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	12/11/2022	12/11/2022
					CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	12/11/2022	12/11/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	12/11/2022	12/11/2022
					CT233	Surgical	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	12/11/2022	12/11/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	12/11/2022	12/11/2022
					CT2406	Implantable	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	12/11/2022	12/11/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	12/11/2022	12/11/2022
					CT2407	Active implantable	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	12/11/2022	12/11/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	12/11/2022	12/11/2022
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	12/11/2022	12/11/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	12/11/2022	12/11/2022
					CT2727	Active implant devices/ancillaries	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	12/11/2022	12/11/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	12/11/2022	12/11/2022
					CT2750	Implantable cardiac electrical stimulators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	12/11/2022	12/11/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	12/11/2022	12/11/2022
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	12/11/2022	12/11/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	12/11/2022	12/11/2022
					CT2985	Active	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	12/11/2022	12/11/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	12/11/2022	12/11/2022
					CT3001	Implantable metal	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	12/11/2022	12/11/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	12/11/2022	12/11/2022
					CT3002	Implantable synthetic polymer	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	12/11/2022	12/11/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	12/11/2022	12/11/2022
					CT3006	Implantable inorganic material	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	12/11/2022	12/11/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	12/11/2022	12/11/2022
					CT314	Electrically powered	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	12/11/2022	12/11/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	12/11/2022	12/11/2022
					CT321	Long-term surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	12/11/2022	12/11/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	12/11/2022	12/11/2022
					CT449	Generators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	12/11/2022	12/11/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	12/11/2022	12/11/2022
					CT981	Single use	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	12/11/2022	12/11/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	12/11/2022	12/11/2022
					CT983	Surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	12/11/2022	12/11/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	12/11/2022	12/11/2022
12022	2023		47265	Dual-chamber implantable pacemaker, rate-responsive	CT141	Electrodes/Leads and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem		
							F1901	Additional Surgery	A24	No Apparent Device Problem		
					CT145	Cardiovascular devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem		
							F1901	Additional Surgery	A24	No Apparent Device Problem		
					CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem		
							F1901	Additional Surgery	A24	No Apparent Device Problem		
					CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem		
							F1901	Additional Surgery	A24	No Apparent Device Problem		
					CT233	Surgical	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem		
							F1901	Additional Surgery	A24	No Apparent Device Problem		
					CT2406	Implantable	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem		
							F1901	Additional Surgery	A24	No Apparent Device Problem		
					CT2407	Active implantable	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem		
							F1901	Additional Surgery	A24	No Apparent Device Problem		
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem		
							F1901	Additional Surgery	A24	No Apparent Device Problem		
					CT2727	Active implant devices/ancillaries	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem		
							F1901	Additional Surgery	A24	No Apparent Device Problem		
					CT2750	Implantable cardiac electrical stimulators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem		
							F1901	Additional Surgery	A24	No Apparent Device Problem		
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem		
							F1901	Additional Surgery	A24	No Apparent Device Problem		
					CT2985	Active	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem		
							F1901	Additional Surgery	A24	No Apparent Device Problem		
					CT3001	Implantable metal	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem		
							F1901	Additional Surgery	A24	No Apparent Device Problem		
					CT3002	Implantable synthetic polymer	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem		
							F1901	Additional Surgery	A24	No Apparent Device Problem		
					CT3006	Implantable inorganic material	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem		
							F1901	Additional Surgery	A24	No Apparent Device Problem		
					CT314	Electrically powered	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem		
							F1901	Additional Surgery	A24	No Apparent Device Problem		
					CT321	Long-term surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem		
							F1901	Additional Surgery	A24	No Apparent Device Problem		
					CT449	Generators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem		

						F1901	Additional Surgery	A24	No Apparent Device Problem			
				CT981	Single use	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem			
						F1901	Additional Surgery	A24	No Apparent Device Problem			
				CT983	Surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem			
						F1901	Additional Surgery	A24	No Apparent Device Problem			
12023	2023	All other reportable incidents	47265	Dual-chamber implantable pacemaker, rate-responsive	CT141	Electrodes/Leads and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	04/01/2023	04/01/2023
							F1901	Additional Surgery	A070504	Premature Discharge of Battery	04/01/2023	04/01/2023
					CT145	Cardiovascular devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	04/01/2023	04/01/2023
							F1901	Additional Surgery	A070504	Premature Discharge of Battery	04/01/2023	04/01/2023
					CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	04/01/2023	04/01/2023
							F1901	Additional Surgery	A070504	Premature Discharge of Battery	04/01/2023	04/01/2023
					CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	04/01/2023	04/01/2023
							F1901	Additional Surgery	A070504	Premature Discharge of Battery	04/01/2023	04/01/2023
					CT233	Surgical	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	04/01/2023	04/01/2023
							F1901	Additional Surgery	A070504	Premature Discharge of Battery	04/01/2023	04/01/2023
					CT2406	Implantable	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	04/01/2023	04/01/2023
							F1901	Additional Surgery	A070504	Premature Discharge of Battery	04/01/2023	04/01/2023
					CT2407	Active implantable	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	04/01/2023	04/01/2023
							F1901	Additional Surgery	A070504	Premature Discharge of Battery	04/01/2023	04/01/2023
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	04/01/2023	04/01/2023
							F1901	Additional Surgery	A070504	Premature Discharge of Battery	04/01/2023	04/01/2023
					CT2727	Active implant devices/ancillaries	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	04/01/2023	04/01/2023
							F1901	Additional Surgery	A070504	Premature Discharge of Battery	04/01/2023	04/01/2023
					CT2750	Implantable cardiac electrical stimulators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	04/01/2023	04/01/2023
							F1901	Additional Surgery	A070504	Premature Discharge of Battery	04/01/2023	04/01/2023
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	04/01/2023	04/01/2023
							F1901	Additional Surgery	A070504	Premature Discharge of Battery	04/01/2023	04/01/2023
					CT2985	Active	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	04/01/2023	04/01/2023
							F1901	Additional Surgery	A070504	Premature Discharge of Battery	04/01/2023	04/01/2023
					CT3001	Implantable metal	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	04/01/2023	04/01/2023
							F1901	Additional Surgery	A070504	Premature Discharge of Battery	04/01/2023	04/01/2023
					CT3002	Implantable synthetic polymer	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	04/01/2023	04/01/2023
							F1901	Additional Surgery	A070504	Premature Discharge of Battery	04/01/2023	04/01/2023
					CT3006	Implantable inorganic material	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	04/01/2023	04/01/2023
							F1901	Additional Surgery	A070504	Premature Discharge of Battery	04/01/2023	04/01/2023
					CT314	Electrically powered	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	04/01/2023	04/01/2023
							F1901	Additional Surgery	A070504	Premature Discharge of Battery	04/01/2023	04/01/2023
					CT321	Long-term surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	04/01/2023	04/01/2023
							F1901	Additional Surgery	A070504	Premature Discharge of Battery	04/01/2023	04/01/2023
					CT449	Generators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	04/01/2023	04/01/2023
							F1901	Additional Surgery	A070504	Premature Discharge of Battery	04/01/2023	04/01/2023
					CT981	Single use	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	04/01/2023	04/01/2023
							F1901	Additional Surgery	A070504	Premature Discharge of Battery	04/01/2023	04/01/2023
					CT983	Surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	04/01/2023	04/01/2023
							F1901	Additional Surgery	A070504	Premature Discharge of Battery	04/01/2023	04/01/2023
12024	2023		47265	Dual-chamber implantable pacemaker, rate-responsive	CT141	Electrodes/Leads and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	04/01/2023	04/01/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	04/01/2023	04/01/2023
					CT145	Cardiovascular devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	04/01/2023	04/01/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	04/01/2023	04/01/2023
					CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	04/01/2023	04/01/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	04/01/2023	04/01/2023
					CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	04/01/2023	04/01/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	04/01/2023	04/01/2023
					CT233	Surgical	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	04/01/2023	04/01/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	04/01/2023	04/01/2023
					CT2406	Implantable	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	04/01/2023	04/01/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	04/01/2023	04/01/2023
					CT2407	Active implantable	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	04/01/2023	04/01/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	04/01/2023	04/01/2023
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	04/01/2023	04/01/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	04/01/2023	04/01/2023
					CT2727	Active implant devices/ancillaries	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	04/01/2023	04/01/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	04/01/2023	04/01/2023
					CT2750	Implantable cardiac electrical stimulators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	04/01/2023	04/01/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	04/01/2023	04/01/2023
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	04/01/2023	04/01/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	04/01/2023	04/01/2023
					CT2985	Active	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	04/01/2023	04/01/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	04/01/2023	04/01/2023
					CT3001	Implantable metal	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	04/01/2023	04/01/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	04/01/2023	04/01/2023
					CT3002	Implantable synthetic polymer	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	04/01/2023	04/01/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	04/01/2023	04/01/2023
					CT3006	Implantable inorganic material	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	04/01/2023	04/01/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	04/01/2023	04/01/2023
					CT314	Electrically powered	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	04/01/2023	04/01/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	04/01/2023	04/01/2023
					CT321	Long-term surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	04/01/2023	04/01/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	04/01/2023	04/01/2023
					CT449	Generators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	04/01/2023	04/01/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	04/01/2023	04/01/2023
					CT981	Single use	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	04/01/2023	04/01/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	04/01/2023	04/01/2023
					CT983	Surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	04/01/2023	04/01/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	04/01/2023	04/01/2023
12025	2023		47265	Dual-chamber implantable pacemaker, rate-responsive	CT141	Electrodes/Leads and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem		
							F1901	Additional Surgery	A24	No Apparent Device Problem		
					CT145	Cardiovascular devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem		

							F1901	Additional Surgery	A24	No Apparent Device Problem		
					CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem		
							F1901	Additional Surgery	A24	No Apparent Device Problem		
					CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem		
							F1901	Additional Surgery	A24	No Apparent Device Problem		
					CT233	Surgical	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem		
							F1901	Additional Surgery	A24	No Apparent Device Problem		
					CT2406	Implantable	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem		
							F1901	Additional Surgery	A24	No Apparent Device Problem		
					CT2407	Active implantable	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem		
							F1901	Additional Surgery	A24	No Apparent Device Problem		
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem		
							F1901	Additional Surgery	A24	No Apparent Device Problem		
					CT2727	Active implant devices/ancillaries	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem		
							F1901	Additional Surgery	A24	No Apparent Device Problem		
					CT2750	Implantable cardiac electrical stimulators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem		
							F1901	Additional Surgery	A24	No Apparent Device Problem		
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem		
							F1901	Additional Surgery	A24	No Apparent Device Problem		
					CT2985	Active	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem		
							F1901	Additional Surgery	A24	No Apparent Device Problem		
					CT3001	Implantable metal	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem		
							F1901	Additional Surgery	A24	No Apparent Device Problem		
					CT3002	Implantable synthetic polymer	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem		
							F1901	Additional Surgery	A24	No Apparent Device Problem		
					CT3006	Implantable inorganic material	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem		
							F1901	Additional Surgery	A24	No Apparent Device Problem		
					CT314	Electrically powered	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem		
							F1901	Additional Surgery	A24	No Apparent Device Problem		
					CT321	Long-term surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem		
							F1901	Additional Surgery	A24	No Apparent Device Problem		
					CT449	Generators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem		
							F1901	Additional Surgery	A24	No Apparent Device Problem		
					CT981	Single use	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem		
							F1901	Additional Surgery	A24	No Apparent Device Problem		
					CT983	Surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem		
							F1901	Additional Surgery	A24	No Apparent Device Problem		
12026	2023	All other reportable incidents	47265	Dual-chamber implantable pacemaker, rate-responsive	CT141	Electrodes/Leads and associated devices	E011903	Syncope/Fainting	A070909	Oversensing	03/01/2023	03/01/2023
									A071204	No Pacing	03/01/2023	03/01/2023
							E0601	Arrhythmia	A070909	Oversensing	03/01/2023	03/01/2023
									A071204	No Pacing	03/01/2023	03/01/2023
							F1901	Additional Surgery	A070909	Oversensing	03/01/2023	03/01/2023
									A071204	No Pacing	03/01/2023	03/01/2023
					CT145	Cardiovascular devices	E011903	Syncope/Fainting	A070909	Oversensing	03/01/2023	03/01/2023
									A071204	No Pacing	03/01/2023	03/01/2023
							E0601	Arrhythmia	A070909	Oversensing	03/01/2023	03/01/2023
									A071204	No Pacing	03/01/2023	03/01/2023
							F1901	Additional Surgery	A070909	Oversensing	03/01/2023	03/01/2023
									A071204	No Pacing	03/01/2023	03/01/2023
					CT1684	Cardiac pacemakers and associated devices	E011903	Syncope/Fainting	A070909	Oversensing	03/01/2023	03/01/2023
									A071204	No Pacing	03/01/2023	03/01/2023
							E0601	Arrhythmia	A070909	Oversensing	03/01/2023	03/01/2023
									A071204	No Pacing	03/01/2023	03/01/2023
							F1901	Additional Surgery	A070909	Oversensing	03/01/2023	03/01/2023
									A071204	No Pacing	03/01/2023	03/01/2023
					CT1685	Cardiac pacemakers	E011903	Syncope/Fainting	A070909	Oversensing	03/01/2023	03/01/2023
									A071204	No Pacing	03/01/2023	03/01/2023
							E0601	Arrhythmia	A070909	Oversensing	03/01/2023	03/01/2023
									A071204	No Pacing	03/01/2023	03/01/2023
							F1901	Additional Surgery	A070909	Oversensing	03/01/2023	03/01/2023
									A071204	No Pacing	03/01/2023	03/01/2023
					CT233	Surgical	E011903	Syncope/Fainting	A070909	Oversensing	03/01/2023	03/01/2023
									A071204	No Pacing	03/01/2023	03/01/2023
							E0601	Arrhythmia	A070909	Oversensing	03/01/2023	03/01/2023
									A071204	No Pacing	03/01/2023	03/01/2023
							F1901	Additional Surgery	A070909	Oversensing	03/01/2023	03/01/2023
									A071204	No Pacing	03/01/2023	03/01/2023
					CT2406	Implantable	E011903	Syncope/Fainting	A070909	Oversensing	03/01/2023	03/01/2023
									A071204	No Pacing	03/01/2023	03/01/2023
							E0601	Arrhythmia	A070909	Oversensing	03/01/2023	03/01/2023
									A071204	No Pacing	03/01/2023	03/01/2023
							F1901	Additional Surgery	A070909	Oversensing	03/01/2023	03/01/2023
									A071204	No Pacing	03/01/2023	03/01/2023
					CT2407	Active implantable	E011903	Syncope/Fainting	A070909	Oversensing	03/01/2023	03/01/2023
									A071204	No Pacing	03/01/2023	03/01/2023
							E0601	Arrhythmia	A070909	Oversensing	03/01/2023	03/01/2023
									A071204	No Pacing	03/01/2023	03/01/2023
							F1901	Additional Surgery	A070909	Oversensing	03/01/2023	03/01/2023
									A071204	No Pacing	03/01/2023	03/01/2023
					CT257	Implantable cardiac pacemakers	E011903	Syncope/Fainting	A070909	Oversensing	03/01/2023	03/01/2023
									A071204	No Pacing	03/01/2023	03/01/2023
							E0601	Arrhythmia	A070909	Oversensing	03/01/2023	03/01/2023
									A071204	No Pacing	03/01/2023	03/01/2023
							F1901	Additional Surgery	A070909	Oversensing	03/01/2023	03/01/2023
									A071204	No Pacing	03/01/2023	03/01/2023
					CT2727	Active implant devices/ancillaries	E011903	Syncope/Fainting	A070909	Oversensing	03/01/2023	03/01/2023
									A071204	No Pacing	03/01/2023	03/01/2023
							E0601	Arrhythmia	A070909	Oversensing	03/01/2023	03/01/2023

									A071204	No Pacing	03/01/2023	03/01/2023
							F1901	Additional Surgery	A070909	Oversensing	03/01/2023	03/01/2023
									A071204	No Pacing	03/01/2023	03/01/2023
					CT2750	Implantable cardiac electrical stimulators and associated devices	E011903	Syncope/Fainting	A070909	Oversensing	03/01/2023	03/01/2023
									A071204	No Pacing	03/01/2023	03/01/2023
							E0601	Arrhythmia	A070909	Oversensing	03/01/2023	03/01/2023
									A071204	No Pacing	03/01/2023	03/01/2023
							F1901	Additional Surgery	A070909	Oversensing	03/01/2023	03/01/2023
									A071204	No Pacing	03/01/2023	03/01/2023
					CT2866	Implantable cardiac electrical stimulators	E011903	Syncope/Fainting	A070909	Oversensing	03/01/2023	03/01/2023
									A071204	No Pacing	03/01/2023	03/01/2023
							E0601	Arrhythmia	A070909	Oversensing	03/01/2023	03/01/2023
									A071204	No Pacing	03/01/2023	03/01/2023
							F1901	Additional Surgery	A070909	Oversensing	03/01/2023	03/01/2023
									A071204	No Pacing	03/01/2023	03/01/2023
					CT2985	Active	E011903	Syncope/Fainting	A070909	Oversensing	03/01/2023	03/01/2023
									A071204	No Pacing	03/01/2023	03/01/2023
							E0601	Arrhythmia	A070909	Oversensing	03/01/2023	03/01/2023
									A071204	No Pacing	03/01/2023	03/01/2023
							F1901	Additional Surgery	A070909	Oversensing	03/01/2023	03/01/2023
									A071204	No Pacing	03/01/2023	03/01/2023
					CT3001	Implantable metal	E011903	Syncope/Fainting	A070909	Oversensing	03/01/2023	03/01/2023
									A071204	No Pacing	03/01/2023	03/01/2023
							E0601	Arrhythmia	A070909	Oversensing	03/01/2023	03/01/2023
									A071204	No Pacing	03/01/2023	03/01/2023
							F1901	Additional Surgery	A070909	Oversensing	03/01/2023	03/01/2023
									A071204	No Pacing	03/01/2023	03/01/2023
					CT3002	Implantable synthetic polymer	E011903	Syncope/Fainting	A070909	Oversensing	03/01/2023	03/01/2023
									A071204	No Pacing	03/01/2023	03/01/2023
							E0601	Arrhythmia	A070909	Oversensing	03/01/2023	03/01/2023
									A071204	No Pacing	03/01/2023	03/01/2023
							F1901	Additional Surgery	A070909	Oversensing	03/01/2023	03/01/2023
									A071204	No Pacing	03/01/2023	03/01/2023
					CT3006	Implantable inorganic material	E011903	Syncope/Fainting	A070909	Oversensing	03/01/2023	03/01/2023
									A071204	No Pacing	03/01/2023	03/01/2023
							E0601	Arrhythmia	A070909	Oversensing	03/01/2023	03/01/2023
									A071204	No Pacing	03/01/2023	03/01/2023
							F1901	Additional Surgery	A070909	Oversensing	03/01/2023	03/01/2023
									A071204	No Pacing	03/01/2023	03/01/2023
					CT314	Electrically powered	E011903	Syncope/Fainting	A070909	Oversensing	03/01/2023	03/01/2023
									A071204	No Pacing	03/01/2023	03/01/2023
							E0601	Arrhythmia	A070909	Oversensing	03/01/2023	03/01/2023
									A071204	No Pacing	03/01/2023	03/01/2023
							F1901	Additional Surgery	A070909	Oversensing	03/01/2023	03/01/2023
									A071204	No Pacing	03/01/2023	03/01/2023
					CT321	Long-term surgical invasive	E011903	Syncope/Fainting	A070909	Oversensing	03/01/2023	03/01/2023
									A071204	No Pacing	03/01/2023	03/01/2023
							E0601	Arrhythmia	A070909	Oversensing	03/01/2023	03/01/2023
									A071204	No Pacing	03/01/2023	03/01/2023
							F1901	Additional Surgery	A070909	Oversensing	03/01/2023	03/01/2023
									A071204	No Pacing	03/01/2023	03/01/2023
					CT449	Generators and associated devices	E011903	Syncope/Fainting	A070909	Oversensing	03/01/2023	03/01/2023
									A071204	No Pacing	03/01/2023	03/01/2023
							E0601	Arrhythmia	A070909	Oversensing	03/01/2023	03/01/2023
									A071204	No Pacing	03/01/2023	03/01/2023
							F1901	Additional Surgery	A070909	Oversensing	03/01/2023	03/01/2023
									A071204	No Pacing	03/01/2023	03/01/2023
					CT981	Single use	E011903	Syncope/Fainting	A070909	Oversensing	03/01/2023	03/01/2023
									A071204	No Pacing	03/01/2023	03/01/2023
							E0601	Arrhythmia	A070909	Oversensing	03/01/2023	03/01/2023
									A071204	No Pacing	03/01/2023	03/01/2023
							F1901	Additional Surgery	A070909	Oversensing	03/01/2023	03/01/2023
									A071204	No Pacing	03/01/2023	03/01/2023
					CT983	Surgical invasive	E011903	Syncope/Fainting	A070909	Oversensing	03/01/2023	03/01/2023
									A071204	No Pacing	03/01/2023	03/01/2023
							E0601	Arrhythmia	A070909	Oversensing	03/01/2023	03/01/2023
									A071204	No Pacing	03/01/2023	03/01/2023
							F1901	Additional Surgery	A070909	Oversensing	03/01/2023	03/01/2023
									A071204	No Pacing	03/01/2023	03/01/2023
12027	2023		47265	Dual-chamber implantable pacemaker, rate-responsive	CT141	Electrodes/Leads and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem		
							F1901	Additional Surgery	A24	No Apparent Device Problem		
					CT145	Cardiovascular devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem		
							F1901	Additional Surgery	A24	No Apparent Device Problem		
					CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem		
							F1901	Additional Surgery	A24	No Apparent Device Problem		
					CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem		
							F1901	Additional Surgery	A24	No Apparent Device Problem		
					CT233	Surgical	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem		
							F1901	Additional Surgery	A24	No Apparent Device Problem		
					CT2406	Implantable	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem		
							F1901	Additional Surgery	A24	No Apparent Device Problem		
					CT2407	Active implantable	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem		
							F1901	Additional Surgery	A24	No Apparent Device Problem		
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem		
							F1901	Additional Surgery	A24	No Apparent Device Problem		
					CT2727	Active implant devices/ancillaries	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem		
							F1901	Additional Surgery	A24	No Apparent Device Problem		
					CT2750	Implantable cardiac electrical stimulators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem		

							F1901	Additional Surgery	A24	No Apparent Device Problem		
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem		
							F1901	Additional Surgery	A24	No Apparent Device Problem		
					CT2985	Active	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem		
							F1901	Additional Surgery	A24	No Apparent Device Problem		
					CT3001	Implantable metal	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem		
							F1901	Additional Surgery	A24	No Apparent Device Problem		
					CT3002	Implantable synthetic polymer	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem		
							F1901	Additional Surgery	A24	No Apparent Device Problem		
					CT3006	Implantable inorganic material	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem		
							F1901	Additional Surgery	A24	No Apparent Device Problem		
					CT314	Electrically powered	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem		
							F1901	Additional Surgery	A24	No Apparent Device Problem		
					CT321	Long-term surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem		
							F1901	Additional Surgery	A24	No Apparent Device Problem		
					CT449	Generators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem		
							F1901	Additional Surgery	A24	No Apparent Device Problem		
					CT981	Single use	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem		
							F1901	Additional Surgery	A24	No Apparent Device Problem		
					CT983	Surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem		
							F1901	Additional Surgery	A24	No Apparent Device Problem		
12028	2023		47265	Dual-chamber implantable pacemaker, rate-responsive	CT141	Electrodes/Leads and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem		
							F1901	Additional Surgery	A24	No Apparent Device Problem		
					CT145	Cardiovascular devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem		
							F1901	Additional Surgery	A24	No Apparent Device Problem		
					CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem		
							F1901	Additional Surgery	A24	No Apparent Device Problem		
					CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem		
							F1901	Additional Surgery	A24	No Apparent Device Problem		
					CT233	Surgical	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem		
							F1901	Additional Surgery	A24	No Apparent Device Problem		
					CT2406	Implantable	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem		
							F1901	Additional Surgery	A24	No Apparent Device Problem		
					CT2407	Active implantable	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem		
							F1901	Additional Surgery	A24	No Apparent Device Problem		
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem		
							F1901	Additional Surgery	A24	No Apparent Device Problem		
					CT2727	Active implant devices/ancillaries	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem		
							F1901	Additional Surgery	A24	No Apparent Device Problem		
					CT2750	Implantable cardiac electrical stimulators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem		
							F1901	Additional Surgery	A24	No Apparent Device Problem		
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem		
							F1901	Additional Surgery	A24	No Apparent Device Problem		
					CT2985	Active	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem		
							F1901	Additional Surgery	A24	No Apparent Device Problem		
					CT3001	Implantable metal	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem		
							F1901	Additional Surgery	A24	No Apparent Device Problem		
					CT3002	Implantable synthetic polymer	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem		
							F1901	Additional Surgery	A24	No Apparent Device Problem		
					CT3006	Implantable inorganic material	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem		
							F1901	Additional Surgery	A24	No Apparent Device Problem		
					CT314	Electrically powered	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem		
							F1901	Additional Surgery	A24	No Apparent Device Problem		
					CT321	Long-term surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem		
							F1901	Additional Surgery	A24	No Apparent Device Problem		
					CT449	Generators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem		
							F1901	Additional Surgery	A24	No Apparent Device Problem		
					CT981	Single use	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem		
							F1901	Additional Surgery	A24	No Apparent Device Problem		
					CT983	Surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem		
							F1901	Additional Surgery	A24	No Apparent Device Problem		
12029	2023	All other reportable incidents	47265	Dual-chamber implantable pacemaker, rate-responsive	CT141	Electrodes/Leads and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem		
							F1901	Additional Surgery	A24	No Apparent Device Problem		
					CT145	Cardiovascular devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem		
							F1901	Additional Surgery	A24	No Apparent Device Problem		
					CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem		
							F1901	Additional Surgery	A24	No Apparent Device Problem		
					CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem		
							F1901	Additional Surgery	A24	No Apparent Device Problem		
					CT233	Surgical	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem		
							F1901	Additional Surgery	A24	No Apparent Device Problem		
					CT2406	Implantable	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem		
							F1901	Additional Surgery	A24	No Apparent Device Problem		
					CT2407	Active implantable	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem		
							F1901	Additional Surgery	A24	No Apparent Device Problem		
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem		
							F1901	Additional Surgery	A24	No Apparent Device Problem		
					CT2727	Active implant devices/ancillaries	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem		
							F1901	Additional Surgery	A24	No Apparent Device Problem		
					CT2750	Implantable cardiac electrical stimulators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem		
							F1901	Additional Surgery	A24	No Apparent Device Problem		
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem		
							F1901	Additional Surgery	A24	No Apparent Device Problem		
					CT2985	Active	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem		
							F1901	Additional Surgery	A24	No Apparent Device Problem		
					CT3001	Implantable metal	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem		
							F1901	Additional Surgery	A24	No Apparent Device Problem		
					CT3002	Implantable synthetic polymer	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem		

						F1901	Additional Surgery	A24	No Apparent Device Problem		
				CT3006	Implantable inorganic material	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem		
						F1901	Additional Surgery	A24	No Apparent Device Problem		
				CT314	Electrically powered	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem		
						F1901	Additional Surgery	A24	No Apparent Device Problem		
				CT321	Long-term surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem		
						F1901	Additional Surgery	A24	No Apparent Device Problem		
				CT449	Generators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem		
						F1901	Additional Surgery	A24	No Apparent Device Problem		
				CT981	Single use	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem		
						F1901	Additional Surgery	A24	No Apparent Device Problem		
				CT983	Surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem		
						F1901	Additional Surgery	A24	No Apparent Device Problem		
12030	2023		47265	Dual-chamber implantable pacemaker, rate-responsive	CT141	Electrodes/Leads and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	17/10/2022	17/10/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	17/10/2022	17/10/2022
					CT145	Cardiovascular devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	17/10/2022	17/10/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	17/10/2022	17/10/2022
					CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	17/10/2022	17/10/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	17/10/2022	17/10/2022
					CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	17/10/2022	17/10/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	17/10/2022	17/10/2022
					CT233	Surgical	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	17/10/2022	17/10/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	17/10/2022	17/10/2022
					CT2406	Implantable	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	17/10/2022	17/10/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	17/10/2022	17/10/2022
					CT2407	Active implantable	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	17/10/2022	17/10/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	17/10/2022	17/10/2022
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	17/10/2022	17/10/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	17/10/2022	17/10/2022
					CT2727	Active implant devices/ancillaries	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	17/10/2022	17/10/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	17/10/2022	17/10/2022
					CT2750	Implantable cardiac electrical stimulators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	17/10/2022	17/10/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	17/10/2022	17/10/2022
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	17/10/2022	17/10/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	17/10/2022	17/10/2022
					CT2985	Active	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	17/10/2022	17/10/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	17/10/2022	17/10/2022
					CT3001	Implantable metal	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	17/10/2022	17/10/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	17/10/2022	17/10/2022
					CT3002	Implantable synthetic polymer	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	17/10/2022	17/10/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	17/10/2022	17/10/2022
					CT3006	Implantable inorganic material	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	17/10/2022	17/10/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	17/10/2022	17/10/2022
					CT314	Electrically powered	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	17/10/2022	17/10/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	17/10/2022	17/10/2022
					CT321	Long-term surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	17/10/2022	17/10/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	17/10/2022	17/10/2022
					CT449	Generators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	17/10/2022	17/10/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	17/10/2022	17/10/2022
					CT981	Single use	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	17/10/2022	17/10/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	17/10/2022	17/10/2022
					CT983	Surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	17/10/2022	17/10/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	17/10/2022	17/10/2022
12031	2023		47265	Dual-chamber implantable pacemaker, rate-responsive	CT141	Electrodes/Leads and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem		
							F1901	Additional Surgery	A24	No Apparent Device Problem		
					CT145	Cardiovascular devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem		
							F1901	Additional Surgery	A24	No Apparent Device Problem		
					CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem		
							F1901	Additional Surgery	A24	No Apparent Device Problem		
					CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem		
							F1901	Additional Surgery	A24	No Apparent Device Problem		
					CT233	Surgical	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem		
							F1901	Additional Surgery	A24	No Apparent Device Problem		
					CT2406	Implantable	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem		
							F1901	Additional Surgery	A24	No Apparent Device Problem		
					CT2407	Active implantable	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem		
							F1901	Additional Surgery	A24	No Apparent Device Problem		
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem		
							F1901	Additional Surgery	A24	No Apparent Device Problem		
					CT2727	Active implant devices/ancillaries	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem		
							F1901	Additional Surgery	A24	No Apparent Device Problem		
					CT2750	Implantable cardiac electrical stimulators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem		
							F1901	Additional Surgery	A24	No Apparent Device Problem		
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem		
							F1901	Additional Surgery	A24	No Apparent Device Problem		
					CT2985	Active	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem		
							F1901	Additional Surgery	A24	No Apparent Device Problem		
					CT3001	Implantable metal	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem		
							F1901	Additional Surgery	A24	No Apparent Device Problem		
					CT3002	Implantable synthetic polymer	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem		
							F1901	Additional Surgery	A24	No Apparent Device Problem		
					CT3006	Implantable inorganic material	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem		
							F1901	Additional Surgery	A24	No Apparent Device Problem		
					CT314	Electrically powered	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem		
							F1901	Additional Surgery	A24	No Apparent Device Problem		
					CT321	Long-term surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem		
							F1901	Additional Surgery	A24	No Apparent Device Problem		
					CT449	Generators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem		

						F1901	Additional Surgery	A24	No Apparent Device Problem			
				CT981	Single use	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem			
						F1901	Additional Surgery	A24	No Apparent Device Problem			
				CT983	Surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem			
						F1901	Additional Surgery	A24	No Apparent Device Problem			
12032	2023	All other reportable incidents	47265	Dual-chamber implantable pacemaker, rate-responsive	CT141	Electrodes/Leads and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	05/12/2022	05/12/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	05/12/2022	05/12/2022
					CT145	Cardiovascular devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	05/12/2022	05/12/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	05/12/2022	05/12/2022
					CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	05/12/2022	05/12/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	05/12/2022	05/12/2022
					CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	05/12/2022	05/12/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	05/12/2022	05/12/2022
					CT233	Surgical	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	05/12/2022	05/12/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	05/12/2022	05/12/2022
					CT2406	Implantable	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	05/12/2022	05/12/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	05/12/2022	05/12/2022
					CT2407	Active implantable	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	05/12/2022	05/12/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	05/12/2022	05/12/2022
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	05/12/2022	05/12/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	05/12/2022	05/12/2022
					CT2727	Active implant devices/ancillaries	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	05/12/2022	05/12/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	05/12/2022	05/12/2022
					CT2750	Implantable cardiac electrical stimulators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	05/12/2022	05/12/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	05/12/2022	05/12/2022
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	05/12/2022	05/12/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	05/12/2022	05/12/2022
					CT2985	Active	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	05/12/2022	05/12/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	05/12/2022	05/12/2022
					CT3001	Implantable metal	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	05/12/2022	05/12/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	05/12/2022	05/12/2022
					CT3002	Implantable synthetic polymer	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	05/12/2022	05/12/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	05/12/2022	05/12/2022
					CT3006	Implantable inorganic material	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	05/12/2022	05/12/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	05/12/2022	05/12/2022
					CT314	Electrically powered	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	05/12/2022	05/12/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	05/12/2022	05/12/2022
					CT321	Long-term surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	05/12/2022	05/12/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	05/12/2022	05/12/2022
					CT449	Generators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	05/12/2022	05/12/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	05/12/2022	05/12/2022
					CT981	Single use	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	05/12/2022	05/12/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	05/12/2022	05/12/2022
					CT983	Surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	05/12/2022	05/12/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	05/12/2022	05/12/2022
12033	2023		47265	Dual-chamber implantable pacemaker, rate-responsive	CT141	Electrodes/Leads and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	05/01/2023	05/01/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	05/01/2023	05/01/2023
					CT145	Cardiovascular devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	05/01/2023	05/01/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	05/01/2023	05/01/2023
					CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	05/01/2023	05/01/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	05/01/2023	05/01/2023
					CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	05/01/2023	05/01/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	05/01/2023	05/01/2023
					CT233	Surgical	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	05/01/2023	05/01/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	05/01/2023	05/01/2023
					CT2406	Implantable	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	05/01/2023	05/01/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	05/01/2023	05/01/2023
					CT2407	Active implantable	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	05/01/2023	05/01/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	05/01/2023	05/01/2023
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	05/01/2023	05/01/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	05/01/2023	05/01/2023
					CT2727	Active implant devices/ancillaries	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	05/01/2023	05/01/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	05/01/2023	05/01/2023
					CT2750	Implantable cardiac electrical stimulators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	05/01/2023	05/01/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	05/01/2023	05/01/2023
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	05/01/2023	05/01/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	05/01/2023	05/01/2023
					CT2985	Active	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	05/01/2023	05/01/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	05/01/2023	05/01/2023
					CT3001	Implantable metal	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	05/01/2023	05/01/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	05/01/2023	05/01/2023
					CT3002	Implantable synthetic polymer	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	05/01/2023	05/01/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	05/01/2023	05/01/2023
					CT3006	Implantable inorganic material	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	05/01/2023	05/01/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	05/01/2023	05/01/2023
					CT314	Electrically powered	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	05/01/2023	05/01/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	05/01/2023	05/01/2023
					CT321	Long-term surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	05/01/2023	05/01/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	05/01/2023	05/01/2023
					CT449	Generators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	05/01/2023	05/01/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	05/01/2023	05/01/2023
					CT981	Single use	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	05/01/2023	05/01/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	05/01/2023	05/01/2023
					CT983	Surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	05/01/2023	05/01/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	05/01/2023	05/01/2023
12034	2023		47265	Dual-chamber implantable pacemaker, rate-responsive	CT141	Electrodes/Leads and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem		
							F1901	Additional Surgery	A24	No Apparent Device Problem		
					CT145	Cardiovascular devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem		

							F1901	Additional Surgery	A24	No Apparent Device Problem		
					CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem		
							F1901	Additional Surgery	A24	No Apparent Device Problem		
					CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem		
							F1901	Additional Surgery	A24	No Apparent Device Problem		
					CT233	Surgical	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem		
							F1901	Additional Surgery	A24	No Apparent Device Problem		
					CT2406	Implantable	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem		
							F1901	Additional Surgery	A24	No Apparent Device Problem		
					CT2407	Active implantable	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem		
							F1901	Additional Surgery	A24	No Apparent Device Problem		
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem		
							F1901	Additional Surgery	A24	No Apparent Device Problem		
					CT2727	Active implant devices/ancillaries	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem		
							F1901	Additional Surgery	A24	No Apparent Device Problem		
					CT2750	Implantable cardiac electrical stimulators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem		
							F1901	Additional Surgery	A24	No Apparent Device Problem		
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem		
							F1901	Additional Surgery	A24	No Apparent Device Problem		
					CT2985	Active	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem		
							F1901	Additional Surgery	A24	No Apparent Device Problem		
					CT3001	Implantable metal	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem		
							F1901	Additional Surgery	A24	No Apparent Device Problem		
					CT3002	Implantable synthetic polymer	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem		
							F1901	Additional Surgery	A24	No Apparent Device Problem		
					CT3006	Implantable inorganic material	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem		
							F1901	Additional Surgery	A24	No Apparent Device Problem		
					CT314	Electrically powered	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem		
							F1901	Additional Surgery	A24	No Apparent Device Problem		
					CT321	Long-term surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem		
							F1901	Additional Surgery	A24	No Apparent Device Problem		
					CT449	Generators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem		
							F1901	Additional Surgery	A24	No Apparent Device Problem		
					CT981	Single use	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem		
							F1901	Additional Surgery	A24	No Apparent Device Problem		
					CT983	Surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem		
							F1901	Additional Surgery	A24	No Apparent Device Problem		
12035	2023		47265	Dual-chamber implantable pacemaker, rate-responsive	CT141	Electrodes/Leads and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	06/01/2023	06/01/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	06/01/2023	06/01/2023
					CT145	Cardiovascular devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	06/01/2023	06/01/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	06/01/2023	06/01/2023
					CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	06/01/2023	06/01/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	06/01/2023	06/01/2023
					CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	06/01/2023	06/01/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	06/01/2023	06/01/2023
					CT233	Surgical	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	06/01/2023	06/01/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	06/01/2023	06/01/2023
					CT2406	Implantable	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	06/01/2023	06/01/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	06/01/2023	06/01/2023
					CT2407	Active implantable	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	06/01/2023	06/01/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	06/01/2023	06/01/2023
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	06/01/2023	06/01/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	06/01/2023	06/01/2023
					CT2727	Active implant devices/ancillaries	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	06/01/2023	06/01/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	06/01/2023	06/01/2023
					CT2750	Implantable cardiac electrical stimulators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	06/01/2023	06/01/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	06/01/2023	06/01/2023
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	06/01/2023	06/01/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	06/01/2023	06/01/2023
					CT2985	Active	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	06/01/2023	06/01/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	06/01/2023	06/01/2023
					CT3001	Implantable metal	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	06/01/2023	06/01/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	06/01/2023	06/01/2023
					CT3002	Implantable synthetic polymer	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	06/01/2023	06/01/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	06/01/2023	06/01/2023
					CT3006	Implantable inorganic material	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	06/01/2023	06/01/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	06/01/2023	06/01/2023
					CT314	Electrically powered	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	06/01/2023	06/01/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	06/01/2023	06/01/2023
					CT321	Long-term surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	06/01/2023	06/01/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	06/01/2023	06/01/2023
					CT449	Generators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	06/01/2023	06/01/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	06/01/2023	06/01/2023
					CT981	Single use	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	06/01/2023	06/01/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	06/01/2023	06/01/2023
					CT983	Surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	06/01/2023	06/01/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	06/01/2023	06/01/2023
12036	2023		47265	Dual-chamber implantable pacemaker, rate-responsive	CT141	Electrodes/Leads and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	09/01/2023	09/01/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	09/01/2023	09/01/2023
					CT145	Cardiovascular devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	09/01/2023	09/01/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	09/01/2023	09/01/2023
					CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	09/01/2023	09/01/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	09/01/2023	09/01/2023
					CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	09/01/2023	09/01/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	09/01/2023	09/01/2023
					CT233	Surgical	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	09/01/2023	09/01/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	09/01/2023	09/01/2023
					CT2406	Implantable	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	09/01/2023	09/01/2023

							F1901	Additional Surgery	A24	No Apparent Device Problem	09/01/2023	09/01/2023
					CT2407	Active implantable	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	09/01/2023	09/01/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	09/01/2023	09/01/2023
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	09/01/2023	09/01/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	09/01/2023	09/01/2023
					CT2727	Active implant devices/ancillaries	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	09/01/2023	09/01/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	09/01/2023	09/01/2023
					CT2750	Implantable cardiac electrical stimulators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	09/01/2023	09/01/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	09/01/2023	09/01/2023
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	09/01/2023	09/01/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	09/01/2023	09/01/2023
					CT2985	Active	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	09/01/2023	09/01/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	09/01/2023	09/01/2023
					CT3001	Implantable metal	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	09/01/2023	09/01/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	09/01/2023	09/01/2023
					CT3002	Implantable synthetic polymer	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	09/01/2023	09/01/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	09/01/2023	09/01/2023
					CT3006	Implantable inorganic material	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	09/01/2023	09/01/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	09/01/2023	09/01/2023
					CT314	Electrically powered	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	09/01/2023	09/01/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	09/01/2023	09/01/2023
					CT321	Long-term surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	09/01/2023	09/01/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	09/01/2023	09/01/2023
					CT449	Generators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	09/01/2023	09/01/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	09/01/2023	09/01/2023
					CT981	Single use	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	09/01/2023	09/01/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	09/01/2023	09/01/2023
					CT983	Surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	09/01/2023	09/01/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	09/01/2023	09/01/2023
12037	2023		47265	Dual-chamber implantable pacemaker, rate-responsive	CT141	Electrodes/Leads and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	11/01/2023	11/01/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	11/01/2023	11/01/2023
					CT145	Cardiovascular devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	11/01/2023	11/01/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	11/01/2023	11/01/2023
					CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	11/01/2023	11/01/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	11/01/2023	11/01/2023
					CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	11/01/2023	11/01/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	11/01/2023	11/01/2023
					CT233	Surgical	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	11/01/2023	11/01/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	11/01/2023	11/01/2023
					CT2406	Implantable	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	11/01/2023	11/01/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	11/01/2023	11/01/2023
					CT2407	Active implantable	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	11/01/2023	11/01/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	11/01/2023	11/01/2023
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	11/01/2023	11/01/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	11/01/2023	11/01/2023
					CT2727	Active implant devices/ancillaries	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	11/01/2023	11/01/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	11/01/2023	11/01/2023
					CT2750	Implantable cardiac electrical stimulators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	11/01/2023	11/01/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	11/01/2023	11/01/2023
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	11/01/2023	11/01/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	11/01/2023	11/01/2023
					CT2985	Active	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	11/01/2023	11/01/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	11/01/2023	11/01/2023
					CT3001	Implantable metal	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	11/01/2023	11/01/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	11/01/2023	11/01/2023
					CT3002	Implantable synthetic polymer	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	11/01/2023	11/01/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	11/01/2023	11/01/2023
					CT3006	Implantable inorganic material	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	11/01/2023	11/01/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	11/01/2023	11/01/2023
					CT314	Electrically powered	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	11/01/2023	11/01/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	11/01/2023	11/01/2023
					CT321	Long-term surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	11/01/2023	11/01/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	11/01/2023	11/01/2023
					CT449	Generators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	11/01/2023	11/01/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	11/01/2023	11/01/2023
					CT981	Single use	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	11/01/2023	11/01/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	11/01/2023	11/01/2023
					CT983	Surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	11/01/2023	11/01/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	11/01/2023	11/01/2023
12038	2023		47265	Dual-chamber implantable pacemaker, rate-responsive	CT141	Electrodes/Leads and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	09/01/2023	09/01/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	09/01/2023	09/01/2023
					CT145	Cardiovascular devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	09/01/2023	09/01/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	09/01/2023	09/01/2023
					CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	09/01/2023	09/01/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	09/01/2023	09/01/2023
					CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	09/01/2023	09/01/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	09/01/2023	09/01/2023
					CT233	Surgical	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	09/01/2023	09/01/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	09/01/2023	09/01/2023
					CT2406	Implantable	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	09/01/2023	09/01/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	09/01/2023	09/01/2023
					CT2407	Active implantable	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	09/01/2023	09/01/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	09/01/2023	09/01/2023
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	09/01/2023	09/01/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	09/01/2023	09/01/2023
					CT2727	Active implant devices/ancillaries	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	09/01/2023	09/01/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	09/01/2023	09/01/2023
					CT2750	Implantable cardiac electrical stimulators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	09/01/2023	09/01/2023

							F1901	Additional Surgery	A24	No Apparent Device Problem	09/01/2023	09/01/2023
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	09/01/2023	09/01/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	09/01/2023	09/01/2023
					CT2985	Active	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	09/01/2023	09/01/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	09/01/2023	09/01/2023
					CT3001	Implantable metal	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	09/01/2023	09/01/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	09/01/2023	09/01/2023
					CT3002	Implantable synthetic polymer	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	09/01/2023	09/01/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	09/01/2023	09/01/2023
					CT3006	Implantable inorganic material	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	09/01/2023	09/01/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	09/01/2023	09/01/2023
					CT314	Electrically powered	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	09/01/2023	09/01/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	09/01/2023	09/01/2023
					CT321	Long-term surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	09/01/2023	09/01/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	09/01/2023	09/01/2023
					CT449	Generators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	09/01/2023	09/01/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	09/01/2023	09/01/2023
					CT981	Single use	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	09/01/2023	09/01/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	09/01/2023	09/01/2023
					CT983	Surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	09/01/2023	09/01/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	09/01/2023	09/01/2023
12039	2023		47265	Dual-chamber implantable pacemaker, rate-responsive	CT141	Electrodes/Leads and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	05/01/2023	05/01/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	05/01/2023	05/01/2023
					CT145	Cardiovascular devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	05/01/2023	05/01/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	05/01/2023	05/01/2023
					CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	05/01/2023	05/01/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	05/01/2023	05/01/2023
					CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	05/01/2023	05/01/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	05/01/2023	05/01/2023
					CT233	Surgical	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	05/01/2023	05/01/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	05/01/2023	05/01/2023
					CT2406	Implantable	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	05/01/2023	05/01/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	05/01/2023	05/01/2023
					CT2407	Active implantable	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	05/01/2023	05/01/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	05/01/2023	05/01/2023
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	05/01/2023	05/01/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	05/01/2023	05/01/2023
					CT2727	Active implant devices/ancillaries	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	05/01/2023	05/01/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	05/01/2023	05/01/2023
					CT2750	Implantable cardiac electrical stimulators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	05/01/2023	05/01/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	05/01/2023	05/01/2023
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	05/01/2023	05/01/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	05/01/2023	05/01/2023
					CT2985	Active	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	05/01/2023	05/01/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	05/01/2023	05/01/2023
					CT3001	Implantable metal	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	05/01/2023	05/01/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	05/01/2023	05/01/2023
					CT3002	Implantable synthetic polymer	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	05/01/2023	05/01/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	05/01/2023	05/01/2023
					CT3006	Implantable inorganic material	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	05/01/2023	05/01/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	05/01/2023	05/01/2023
					CT314	Electrically powered	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	05/01/2023	05/01/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	05/01/2023	05/01/2023
					CT321	Long-term surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	05/01/2023	05/01/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	05/01/2023	05/01/2023
					CT449	Generators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	05/01/2023	05/01/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	05/01/2023	05/01/2023
					CT981	Single use	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	05/01/2023	05/01/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	05/01/2023	05/01/2023
					CT983	Surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	05/01/2023	05/01/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	05/01/2023	05/01/2023
12040	2023		47265	Dual-chamber implantable pacemaker, rate-responsive	CT141	Electrodes/Leads and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	05/01/2023	05/01/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	05/01/2023	05/01/2023
					CT145	Cardiovascular devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	05/01/2023	05/01/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	05/01/2023	05/01/2023
					CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	05/01/2023	05/01/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	05/01/2023	05/01/2023
					CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	05/01/2023	05/01/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	05/01/2023	05/01/2023
					CT233	Surgical	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	05/01/2023	05/01/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	05/01/2023	05/01/2023
					CT2406	Implantable	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	05/01/2023	05/01/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	05/01/2023	05/01/2023
					CT2407	Active implantable	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	05/01/2023	05/01/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	05/01/2023	05/01/2023
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	05/01/2023	05/01/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	05/01/2023	05/01/2023
					CT2727	Active implant devices/ancillaries	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	05/01/2023	05/01/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	05/01/2023	05/01/2023
					CT2750	Implantable cardiac electrical stimulators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	05/01/2023	05/01/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	05/01/2023	05/01/2023
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	05/01/2023	05/01/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	05/01/2023	05/01/2023
					CT2985	Active	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	05/01/2023	05/01/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	05/01/2023	05/01/2023
					CT3001	Implantable metal	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	05/01/2023	05/01/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	05/01/2023	05/01/2023
					CT3002	Implantable synthetic polymer	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	05/01/2023	05/01/2023

						F1901	Additional Surgery	A24	No Apparent Device Problem	05/01/2023	05/01/2023
				CT3006	Implantable inorganic material	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	05/01/2023	05/01/2023
						F1901	Additional Surgery	A24	No Apparent Device Problem	05/01/2023	05/01/2023
				CT314	Electrically powered	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	05/01/2023	05/01/2023
						F1901	Additional Surgery	A24	No Apparent Device Problem	05/01/2023	05/01/2023
				CT321	Long-term surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	05/01/2023	05/01/2023
						F1901	Additional Surgery	A24	No Apparent Device Problem	05/01/2023	05/01/2023
				CT449	Generators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	05/01/2023	05/01/2023
						F1901	Additional Surgery	A24	No Apparent Device Problem	05/01/2023	05/01/2023
				CT981	Single use	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	05/01/2023	05/01/2023
						F1901	Additional Surgery	A24	No Apparent Device Problem	05/01/2023	05/01/2023
				CT983	Surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	05/01/2023	05/01/2023
						F1901	Additional Surgery	A24	No Apparent Device Problem	05/01/2023	05/01/2023
12041	2023		47265	Dual-chamber implantable pacemaker, rate-responsive	CT141	Electrodes/Leads and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	22/12/2022	22/12/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	22/12/2022	22/12/2022
					CT145	Cardiovascular devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	22/12/2022	22/12/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	22/12/2022	22/12/2022
					CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	22/12/2022	22/12/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	22/12/2022	22/12/2022
					CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	22/12/2022	22/12/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	22/12/2022	22/12/2022
					CT233	Surgical	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	22/12/2022	22/12/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	22/12/2022	22/12/2022
					CT2406	Implantable	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	22/12/2022	22/12/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	22/12/2022	22/12/2022
					CT2407	Active implantable	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	22/12/2022	22/12/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	22/12/2022	22/12/2022
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	22/12/2022	22/12/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	22/12/2022	22/12/2022
					CT2727	Active implant devices/ancillaries	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	22/12/2022	22/12/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	22/12/2022	22/12/2022
					CT2750	Implantable cardiac electrical stimulators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	22/12/2022	22/12/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	22/12/2022	22/12/2022
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	22/12/2022	22/12/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	22/12/2022	22/12/2022
					CT2985	Active	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	22/12/2022	22/12/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	22/12/2022	22/12/2022
					CT3001	Implantable metal	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	22/12/2022	22/12/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	22/12/2022	22/12/2022
					CT3002	Implantable synthetic polymer	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	22/12/2022	22/12/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	22/12/2022	22/12/2022
					CT3006	Implantable inorganic material	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	22/12/2022	22/12/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	22/12/2022	22/12/2022
					CT314	Electrically powered	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	22/12/2022	22/12/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	22/12/2022	22/12/2022
					CT321	Long-term surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	22/12/2022	22/12/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	22/12/2022	22/12/2022
					CT449	Generators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	22/12/2022	22/12/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	22/12/2022	22/12/2022
					CT981	Single use	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	22/12/2022	22/12/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	22/12/2022	22/12/2022
					CT983	Surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	22/12/2022	22/12/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	22/12/2022	22/12/2022
12042	2023		47265	Dual-chamber implantable pacemaker, rate-responsive	CT141	Electrodes/Leads and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	09/11/2022	09/11/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	09/11/2022	09/11/2022
					CT145	Cardiovascular devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	09/11/2022	09/11/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	09/11/2022	09/11/2022
					CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	09/11/2022	09/11/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	09/11/2022	09/11/2022
					CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	09/11/2022	09/11/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	09/11/2022	09/11/2022
					CT233	Surgical	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	09/11/2022	09/11/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	09/11/2022	09/11/2022
					CT2406	Implantable	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	09/11/2022	09/11/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	09/11/2022	09/11/2022
					CT2407	Active implantable	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	09/11/2022	09/11/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	09/11/2022	09/11/2022
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	09/11/2022	09/11/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	09/11/2022	09/11/2022
					CT2727	Active implant devices/ancillaries	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	09/11/2022	09/11/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	09/11/2022	09/11/2022
					CT2750	Implantable cardiac electrical stimulators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	09/11/2022	09/11/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	09/11/2022	09/11/2022
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	09/11/2022	09/11/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	09/11/2022	09/11/2022
					CT2985	Active	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	09/11/2022	09/11/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	09/11/2022	09/11/2022
					CT3001	Implantable metal	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	09/11/2022	09/11/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	09/11/2022	09/11/2022
					CT3002	Implantable synthetic polymer	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	09/11/2022	09/11/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	09/11/2022	09/11/2022
					CT3006	Implantable inorganic material	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	09/11/2022	09/11/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	09/11/2022	09/11/2022
					CT314	Electrically powered	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	09/11/2022	09/11/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	09/11/2022	09/11/2022
					CT321	Long-term surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	09/11/2022	09/11/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	09/11/2022	09/11/2022
					CT449	Generators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	09/11/2022	09/11/2022

						F1901	Additional Surgery	A24	No Apparent Device Problem	09/11/2022	09/11/2022
				CT981	Single use	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	09/11/2022	09/11/2022
						F1901	Additional Surgery	A24	No Apparent Device Problem	09/11/2022	09/11/2022
				CT983	Surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	09/11/2022	09/11/2022
						F1901	Additional Surgery	A24	No Apparent Device Problem	09/11/2022	09/11/2022
12043	2023	All other reportable incidents	47265	Dual-chamber implantable pacemaker, rate-responsive	CT141	Electrodes/Leads and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070909	Oversensing	16/11/2022	16/11/2022
							F1901	Additional Surgery	A070909	Oversensing	16/11/2022	16/11/2022
					CT145	Cardiovascular devices	E2403	No Clinical Signs, Symptoms or Conditions	A070909	Oversensing	16/11/2022	16/11/2022
							F1901	Additional Surgery	A070909	Oversensing	16/11/2022	16/11/2022
					CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070909	Oversensing	16/11/2022	16/11/2022
							F1901	Additional Surgery	A070909	Oversensing	16/11/2022	16/11/2022
					CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A070909	Oversensing	16/11/2022	16/11/2022
							F1901	Additional Surgery	A070909	Oversensing	16/11/2022	16/11/2022
					CT233	Surgical	E2403	No Clinical Signs, Symptoms or Conditions	A070909	Oversensing	16/11/2022	16/11/2022
							F1901	Additional Surgery	A070909	Oversensing	16/11/2022	16/11/2022
					CT2406	Implantable	E2403	No Clinical Signs, Symptoms or Conditions	A070909	Oversensing	16/11/2022	16/11/2022
							F1901	Additional Surgery	A070909	Oversensing	16/11/2022	16/11/2022
					CT2407	Active implantable	E2403	No Clinical Signs, Symptoms or Conditions	A070909	Oversensing	16/11/2022	16/11/2022
							F1901	Additional Surgery	A070909	Oversensing	16/11/2022	16/11/2022
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A070909	Oversensing	16/11/2022	16/11/2022
							F1901	Additional Surgery	A070909	Oversensing	16/11/2022	16/11/2022
					CT2727	Active implant devices/ancillaries	E2403	No Clinical Signs, Symptoms or Conditions	A070909	Oversensing	16/11/2022	16/11/2022
							F1901	Additional Surgery	A070909	Oversensing	16/11/2022	16/11/2022
					CT2750	Implantable cardiac electrical stimulators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070909	Oversensing	16/11/2022	16/11/2022
							F1901	Additional Surgery	A070909	Oversensing	16/11/2022	16/11/2022
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A070909	Oversensing	16/11/2022	16/11/2022
							F1901	Additional Surgery	A070909	Oversensing	16/11/2022	16/11/2022
					CT2985	Active	E2403	No Clinical Signs, Symptoms or Conditions	A070909	Oversensing	16/11/2022	16/11/2022
							F1901	Additional Surgery	A070909	Oversensing	16/11/2022	16/11/2022
					CT3001	Implantable metal	E2403	No Clinical Signs, Symptoms or Conditions	A070909	Oversensing	16/11/2022	16/11/2022
							F1901	Additional Surgery	A070909	Oversensing	16/11/2022	16/11/2022
					CT3002	Implantable synthetic polymer	E2403	No Clinical Signs, Symptoms or Conditions	A070909	Oversensing	16/11/2022	16/11/2022
							F1901	Additional Surgery	A070909	Oversensing	16/11/2022	16/11/2022
					CT3006	Implantable inorganic material	E2403	No Clinical Signs, Symptoms or Conditions	A070909	Oversensing	16/11/2022	16/11/2022
							F1901	Additional Surgery	A070909	Oversensing	16/11/2022	16/11/2022
					CT314	Electrically powered	E2403	No Clinical Signs, Symptoms or Conditions	A070909	Oversensing	16/11/2022	16/11/2022
							F1901	Additional Surgery	A070909	Oversensing	16/11/2022	16/11/2022
					CT321	Long-term surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A070909	Oversensing	16/11/2022	16/11/2022
							F1901	Additional Surgery	A070909	Oversensing	16/11/2022	16/11/2022
					CT449	Generators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070909	Oversensing	16/11/2022	16/11/2022
							F1901	Additional Surgery	A070909	Oversensing	16/11/2022	16/11/2022
					CT981	Single use	E2403	No Clinical Signs, Symptoms or Conditions	A070909	Oversensing	16/11/2022	16/11/2022
							F1901	Additional Surgery	A070909	Oversensing	16/11/2022	16/11/2022
					CT983	Surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A070909	Oversensing	16/11/2022	16/11/2022
							F1901	Additional Surgery	A070909	Oversensing	16/11/2022	16/11/2022
12044	2023		47265	Dual-chamber implantable pacemaker, rate-responsive	CT141	Electrodes/Leads and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	06/01/2023	06/01/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	06/01/2023	06/01/2023
					CT145	Cardiovascular devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	06/01/2023	06/01/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	06/01/2023	06/01/2023
					CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	06/01/2023	06/01/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	06/01/2023	06/01/2023
					CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	06/01/2023	06/01/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	06/01/2023	06/01/2023
					CT233	Surgical	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	06/01/2023	06/01/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	06/01/2023	06/01/2023
					CT2406	Implantable	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	06/01/2023	06/01/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	06/01/2023	06/01/2023
					CT2407	Active implantable	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	06/01/2023	06/01/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	06/01/2023	06/01/2023
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	06/01/2023	06/01/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	06/01/2023	06/01/2023
					CT2727	Active implant devices/ancillaries	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	06/01/2023	06/01/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	06/01/2023	06/01/2023
					CT2750	Implantable cardiac electrical stimulators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	06/01/2023	06/01/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	06/01/2023	06/01/2023
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	06/01/2023	06/01/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	06/01/2023	06/01/2023
					CT2985	Active	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	06/01/2023	06/01/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	06/01/2023	06/01/2023
					CT3001	Implantable metal	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	06/01/2023	06/01/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	06/01/2023	06/01/2023
					CT3002	Implantable synthetic polymer	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	06/01/2023	06/01/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	06/01/2023	06/01/2023
					CT3006	Implantable inorganic material	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	06/01/2023	06/01/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	06/01/2023	06/01/2023
					CT314	Electrically powered	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	06/01/2023	06/01/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	06/01/2023	06/01/2023
					CT321	Long-term surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	06/01/2023	06/01/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	06/01/2023	06/01/2023
					CT449	Generators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	06/01/2023	06/01/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	06/01/2023	06/01/2023
					CT981	Single use	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	06/01/2023	06/01/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	06/01/2023	06/01/2023
					CT983	Surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	06/01/2023	06/01/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	06/01/2023	06/01/2023
12045	2023	All other reportable incidents	17882	Physiologic-monitoring defibrillation system	CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A0719	Unexpected Shutdown		
							F26	No Health Consequences or Impact	A0719	Unexpected Shutdown		
12046	2023		47265	Dual-chamber implantable pacemaker, rate-	CT141	Electrodes/Leads and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	05/01/2023	05/01/2023

				responsive			F1901	Additional Surgery	A24	No Apparent Device Problem	05/01/2023	05/01/2023
					CT145	Cardiovascular devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	05/01/2023	05/01/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	05/01/2023	05/01/2023
					CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	05/01/2023	05/01/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	05/01/2023	05/01/2023
					CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	05/01/2023	05/01/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	05/01/2023	05/01/2023
					CT233	Surgical	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	05/01/2023	05/01/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	05/01/2023	05/01/2023
					CT2406	Implantable	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	05/01/2023	05/01/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	05/01/2023	05/01/2023
					CT2407	Active implantable	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	05/01/2023	05/01/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	05/01/2023	05/01/2023
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	05/01/2023	05/01/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	05/01/2023	05/01/2023
					CT2727	Active implant devices/ancillaries	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	05/01/2023	05/01/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	05/01/2023	05/01/2023
					CT2750	Implantable cardiac electrical stimulators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	05/01/2023	05/01/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	05/01/2023	05/01/2023
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	05/01/2023	05/01/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	05/01/2023	05/01/2023
					CT2985	Active	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	05/01/2023	05/01/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	05/01/2023	05/01/2023
					CT3001	Implantable metal	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	05/01/2023	05/01/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	05/01/2023	05/01/2023
					CT3002	Implantable synthetic polymer	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	05/01/2023	05/01/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	05/01/2023	05/01/2023
					CT3006	Implantable inorganic material	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	05/01/2023	05/01/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	05/01/2023	05/01/2023
					CT314	Electrically powered	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	05/01/2023	05/01/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	05/01/2023	05/01/2023
					CT321	Long-term surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	05/01/2023	05/01/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	05/01/2023	05/01/2023
					CT449	Generators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	05/01/2023	05/01/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	05/01/2023	05/01/2023
					CT981	Single use	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	05/01/2023	05/01/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	05/01/2023	05/01/2023
					CT983	Surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	05/01/2023	05/01/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	05/01/2023	05/01/2023
12047	2023		47265	Dual-chamber implantable pacemaker, rate-responsive	CT141	Electrodes/Leads and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	05/01/2023	05/01/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	05/01/2023	05/01/2023
					CT145	Cardiovascular devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	05/01/2023	05/01/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	05/01/2023	05/01/2023
					CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	05/01/2023	05/01/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	05/01/2023	05/01/2023
					CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	05/01/2023	05/01/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	05/01/2023	05/01/2023
					CT233	Surgical	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	05/01/2023	05/01/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	05/01/2023	05/01/2023
					CT2406	Implantable	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	05/01/2023	05/01/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	05/01/2023	05/01/2023
					CT2407	Active implantable	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	05/01/2023	05/01/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	05/01/2023	05/01/2023
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	05/01/2023	05/01/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	05/01/2023	05/01/2023
					CT2727	Active implant devices/ancillaries	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	05/01/2023	05/01/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	05/01/2023	05/01/2023
					CT2750	Implantable cardiac electrical stimulators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	05/01/2023	05/01/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	05/01/2023	05/01/2023
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	05/01/2023	05/01/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	05/01/2023	05/01/2023
					CT2985	Active	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	05/01/2023	05/01/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	05/01/2023	05/01/2023
					CT3001	Implantable metal	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	05/01/2023	05/01/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	05/01/2023	05/01/2023
					CT3002	Implantable synthetic polymer	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	05/01/2023	05/01/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	05/01/2023	05/01/2023
					CT3006	Implantable inorganic material	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	05/01/2023	05/01/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	05/01/2023	05/01/2023
					CT314	Electrically powered	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	05/01/2023	05/01/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	05/01/2023	05/01/2023
					CT321	Long-term surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	05/01/2023	05/01/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	05/01/2023	05/01/2023
					CT449	Generators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	05/01/2023	05/01/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	05/01/2023	05/01/2023
					CT981	Single use	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	05/01/2023	05/01/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	05/01/2023	05/01/2023
					CT983	Surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	05/01/2023	05/01/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	05/01/2023	05/01/2023
12048	2023	All other reportable incidents	37265	Dual-chamber implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery		
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery		
12049	2023		47265	Dual-chamber implantable pacemaker, rate-responsive	CT141	Electrodes/Leads and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem		
							F1901	Additional Surgery	A24	No Apparent Device Problem		
					CT145	Cardiovascular devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem		
							F1901	Additional Surgery	A24	No Apparent Device Problem		
					CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem		
							F1901	Additional Surgery	A24	No Apparent Device Problem		

					CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem		
							F1901	Additional Surgery	A24	No Apparent Device Problem		
					CT233	Surgical	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem		
							F1901	Additional Surgery	A24	No Apparent Device Problem		
					CT2406	Implantable	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem		
							F1901	Additional Surgery	A24	No Apparent Device Problem		
					CT2407	Active implantable	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem		
							F1901	Additional Surgery	A24	No Apparent Device Problem		
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem		
							F1901	Additional Surgery	A24	No Apparent Device Problem		
					CT2727	Active implant devices/ancillaries	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem		
							F1901	Additional Surgery	A24	No Apparent Device Problem		
					CT2750	Implantable cardiac electrical stimulators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem		
							F1901	Additional Surgery	A24	No Apparent Device Problem		
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem		
							F1901	Additional Surgery	A24	No Apparent Device Problem		
					CT2985	Active	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem		
							F1901	Additional Surgery	A24	No Apparent Device Problem		
					CT3001	Implantable metal	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem		
							F1901	Additional Surgery	A24	No Apparent Device Problem		
					CT3002	Implantable synthetic polymer	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem		
							F1901	Additional Surgery	A24	No Apparent Device Problem		
					CT3006	Implantable inorganic material	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem		
							F1901	Additional Surgery	A24	No Apparent Device Problem		
					CT314	Electrically powered	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem		
							F1901	Additional Surgery	A24	No Apparent Device Problem		
					CT321	Long-term surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem		
							F1901	Additional Surgery	A24	No Apparent Device Problem		
					CT449	Generators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem		
							F1901	Additional Surgery	A24	No Apparent Device Problem		
					CT981	Single use	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem		
							F1901	Additional Surgery	A24	No Apparent Device Problem		
					CT983	Surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem		
							F1901	Additional Surgery	A24	No Apparent Device Problem		
12050	2023	All other reportable incidents	35223	Endocardial/interventricular septal pacing lead	CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A0401	Break	05/01/2023	05/01/2023
									A070909	Oversensing	05/01/2023	05/01/2023
							F1901	Additional Surgery	A0401	Break	05/01/2023	05/01/2023
									A070909	Oversensing	05/01/2023	05/01/2023
12051	2023		47265	Dual-chamber implantable pacemaker, rate-responsive	CT141	Electrodes/Leads and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem		
							F1901	Additional Surgery	A24	No Apparent Device Problem		
					CT145	Cardiovascular devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem		
							F1901	Additional Surgery	A24	No Apparent Device Problem		
					CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem		
							F1901	Additional Surgery	A24	No Apparent Device Problem		
					CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem		
							F1901	Additional Surgery	A24	No Apparent Device Problem		
					CT233	Surgical	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem		
							F1901	Additional Surgery	A24	No Apparent Device Problem		
					CT2406	Implantable	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem		
							F1901	Additional Surgery	A24	No Apparent Device Problem		
					CT2407	Active implantable	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem		
							F1901	Additional Surgery	A24	No Apparent Device Problem		
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem		
							F1901	Additional Surgery	A24	No Apparent Device Problem		
					CT2727	Active implant devices/ancillaries	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem		
							F1901	Additional Surgery	A24	No Apparent Device Problem		
					CT2750	Implantable cardiac electrical stimulators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem		
							F1901	Additional Surgery	A24	No Apparent Device Problem		
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem		
							F1901	Additional Surgery	A24	No Apparent Device Problem		
					CT2985	Active	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem		
							F1901	Additional Surgery	A24	No Apparent Device Problem		
					CT3001	Implantable metal	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem		
							F1901	Additional Surgery	A24	No Apparent Device Problem		
					CT3002	Implantable synthetic polymer	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem		
							F1901	Additional Surgery	A24	No Apparent Device Problem		
					CT3006	Implantable inorganic material	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem		
							F1901	Additional Surgery	A24	No Apparent Device Problem		
					CT314	Electrically powered	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem		
							F1901	Additional Surgery	A24	No Apparent Device Problem		
					CT321	Long-term surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem		
							F1901	Additional Surgery	A24	No Apparent Device Problem		
					CT449	Generators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem		
							F1901	Additional Surgery	A24	No Apparent Device Problem		
					CT981	Single use	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem		
							F1901	Additional Surgery	A24	No Apparent Device Problem		
					CT983	Surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem		
							F1901	Additional Surgery	A24	No Apparent Device Problem		
12052	2023		47265	Dual-chamber implantable pacemaker, rate-responsive	CT141	Electrodes/Leads and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem		
							F1901	Additional Surgery	A24	No Apparent Device Problem		
					CT145	Cardiovascular devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem		
							F1901	Additional Surgery	A24	No Apparent Device Problem		
					CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem		
							F1901	Additional Surgery	A24	No Apparent Device Problem		
					CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem		
							F1901	Additional Surgery	A24	No Apparent Device Problem		
					CT233	Surgical	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem		
							F1901	Additional Surgery	A24	No Apparent Device Problem		

					CT2406	Implantable	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem		
							F1901	Additional Surgery	A24	No Apparent Device Problem		
					CT2407	Active implantable	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem		
							F1901	Additional Surgery	A24	No Apparent Device Problem		
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem		
							F1901	Additional Surgery	A24	No Apparent Device Problem		
					CT2727	Active implant devices/ancillaries	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem		
							F1901	Additional Surgery	A24	No Apparent Device Problem		
					CT2750	Implantable cardiac electrical stimulators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem		
							F1901	Additional Surgery	A24	No Apparent Device Problem		
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem		
							F1901	Additional Surgery	A24	No Apparent Device Problem		
					CT2985	Active	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem		
							F1901	Additional Surgery	A24	No Apparent Device Problem		
					CT3001	Implantable metal	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem		
							F1901	Additional Surgery	A24	No Apparent Device Problem		
					CT3002	Implantable synthetic polymer	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem		
							F1901	Additional Surgery	A24	No Apparent Device Problem		
					CT3006	Implantable inorganic material	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem		
							F1901	Additional Surgery	A24	No Apparent Device Problem		
					CT314	Electrically powered	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem		
							F1901	Additional Surgery	A24	No Apparent Device Problem		
					CT321	Long-term surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem		
							F1901	Additional Surgery	A24	No Apparent Device Problem		
					CT449	Generators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem		
							F1901	Additional Surgery	A24	No Apparent Device Problem		
					CT981	Single use	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem		
							F1901	Additional Surgery	A24	No Apparent Device Problem		
					CT983	Surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem		
							F1901	Additional Surgery	A24	No Apparent Device Problem		
12053	2023		47265	Dual-chamber implantable pacemaker, rate-responsive	CT141	Electrodes/Leads and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem		
							F1901	Additional Surgery	A24	No Apparent Device Problem		
					CT145	Cardiovascular devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem		
							F1901	Additional Surgery	A24	No Apparent Device Problem		
					CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem		
							F1901	Additional Surgery	A24	No Apparent Device Problem		
					CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem		
							F1901	Additional Surgery	A24	No Apparent Device Problem		
					CT233	Surgical	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem		
							F1901	Additional Surgery	A24	No Apparent Device Problem		
					CT2406	Implantable	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem		
							F1901	Additional Surgery	A24	No Apparent Device Problem		
					CT2407	Active implantable	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem		
							F1901	Additional Surgery	A24	No Apparent Device Problem		
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem		
							F1901	Additional Surgery	A24	No Apparent Device Problem		
					CT2727	Active implant devices/ancillaries	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem		
							F1901	Additional Surgery	A24	No Apparent Device Problem		
					CT2750	Implantable cardiac electrical stimulators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem		
							F1901	Additional Surgery	A24	No Apparent Device Problem		
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem		
							F1901	Additional Surgery	A24	No Apparent Device Problem		
					CT2985	Active	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem		
							F1901	Additional Surgery	A24	No Apparent Device Problem		
					CT3001	Implantable metal	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem		
							F1901	Additional Surgery	A24	No Apparent Device Problem		
					CT3002	Implantable synthetic polymer	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem		
							F1901	Additional Surgery	A24	No Apparent Device Problem		
					CT3006	Implantable inorganic material	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem		
							F1901	Additional Surgery	A24	No Apparent Device Problem		
					CT314	Electrically powered	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem		
							F1901	Additional Surgery	A24	No Apparent Device Problem		
					CT321	Long-term surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem		
							F1901	Additional Surgery	A24	No Apparent Device Problem		
					CT449	Generators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem		
							F1901	Additional Surgery	A24	No Apparent Device Problem		
					CT981	Single use	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem		
							F1901	Additional Surgery	A24	No Apparent Device Problem		
					CT983	Surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem		
							F1901	Additional Surgery	A24	No Apparent Device Problem		
12054	2023	All other reportable incidents	47265	Dual-chamber implantable pacemaker, rate-responsive	CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
12055	2023	All other reportable incidents	35223	Endocardial/interventricular septal pacing lead	CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070909	Oversensing	22/09/2022	22/09/2022
							F1901	Additional Surgery	A070909	Oversensing	22/09/2022	22/09/2022
12056	2023		47265	Dual-chamber implantable pacemaker, rate-responsive	CT141	Electrodes/Leads and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	22/11/2022	22/11/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	22/11/2022	22/11/2022
					CT145	Cardiovascular devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	22/11/2022	22/11/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	22/11/2022	22/11/2022

					CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	22/11/2022	22/11/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	22/11/2022	22/11/2022
					CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	22/11/2022	22/11/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	22/11/2022	22/11/2022
					CT233	Surgical	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	22/11/2022	22/11/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	22/11/2022	22/11/2022
					CT2406	Implantable	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	22/11/2022	22/11/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	22/11/2022	22/11/2022
					CT2407	Active implantable	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	22/11/2022	22/11/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	22/11/2022	22/11/2022
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	22/11/2022	22/11/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	22/11/2022	22/11/2022
					CT2727	Active implant devices/ancillaries	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	22/11/2022	22/11/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	22/11/2022	22/11/2022
					CT2750	Implantable cardiac electrical stimulators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	22/11/2022	22/11/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	22/11/2022	22/11/2022
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	22/11/2022	22/11/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	22/11/2022	22/11/2022
					CT2985	Active	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	22/11/2022	22/11/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	22/11/2022	22/11/2022
					CT3001	Implantable metal	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	22/11/2022	22/11/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	22/11/2022	22/11/2022
					CT3002	Implantable synthetic polymer	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	22/11/2022	22/11/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	22/11/2022	22/11/2022
					CT3006	Implantable inorganic material	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	22/11/2022	22/11/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	22/11/2022	22/11/2022
					CT314	Electrically powered	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	22/11/2022	22/11/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	22/11/2022	22/11/2022
					CT321	Long-term surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	22/11/2022	22/11/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	22/11/2022	22/11/2022
					CT449	Generators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	22/11/2022	22/11/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	22/11/2022	22/11/2022
					CT981	Single use	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	22/11/2022	22/11/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	22/11/2022	22/11/2022
					CT983	Surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	22/11/2022	22/11/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	22/11/2022	22/11/2022
12057	2023	All other reportable incidents	35223	Endocardial/interventricular septal pacing lead	CT1684	Cardiac pacemakers and associated devices	E0112	Dizziness	A0401	Break		
									A070909	Oversensing		
									A0722	Impedance Problem		
							E0601	Arrhythmia	A0401	Break		
									A070909	Oversensing		
									A0722	Impedance Problem		
							E1720	Skin Inflammation/ Irritation	A0401	Break		
									A070909	Oversensing		
									A0722	Impedance Problem		
							F11	Non-Serious Injury/ Illness/ Impairment	A0401	Break		
									A070909	Oversensing		
									A0722	Impedance Problem		
12058	2023		47265	Dual-chamber implantable pacemaker, rate-responsive	CT141	Electrodes/Leads and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	12/01/2023	12/01/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	12/01/2023	12/01/2023
					CT145	Cardiovascular devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	12/01/2023	12/01/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	12/01/2023	12/01/2023
					CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	12/01/2023	12/01/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	12/01/2023	12/01/2023
					CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	12/01/2023	12/01/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	12/01/2023	12/01/2023
					CT233	Surgical	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	12/01/2023	12/01/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	12/01/2023	12/01/2023
					CT2406	Implantable	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	12/01/2023	12/01/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	12/01/2023	12/01/2023
					CT2407	Active implantable	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	12/01/2023	12/01/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	12/01/2023	12/01/2023
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	12/01/2023	12/01/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	12/01/2023	12/01/2023
					CT2727	Active implant devices/ancillaries	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	12/01/2023	12/01/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	12/01/2023	12/01/2023
					CT2750	Implantable cardiac electrical stimulators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	12/01/2023	12/01/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	12/01/2023	12/01/2023
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	12/01/2023	12/01/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	12/01/2023	12/01/2023
					CT2985	Active	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	12/01/2023	12/01/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	12/01/2023	12/01/2023
					CT3001	Implantable metal	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	12/01/2023	12/01/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	12/01/2023	12/01/2023
					CT3002	Implantable synthetic polymer	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	12/01/2023	12/01/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	12/01/2023	12/01/2023
					CT3006	Implantable inorganic material	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	12/01/2023	12/01/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	12/01/2023	12/01/2023
					CT314	Electrically powered	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	12/01/2023	12/01/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	12/01/2023	12/01/2023
					CT321	Long-term surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	12/01/2023	12/01/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	12/01/2023	12/01/2023
					CT449	Generators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	12/01/2023	12/01/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	12/01/2023	12/01/2023
					CT981	Single use	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	12/01/2023	12/01/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	12/01/2023	12/01/2023
					CT983	Surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	12/01/2023	12/01/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	12/01/2023	12/01/2023

12059	2023		47265	Dual-chamber implantable pacemaker, rate-responsive	CT141	Electrodes/Leads and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A071204	No Pacing	30/12/2022	30/12/2022
							F1908	Prolonged Surgery	A071204	No Pacing	30/12/2022	30/12/2022
					CT145	Cardiovascular devices	E2403	No Clinical Signs, Symptoms or Conditions	A071204	No Pacing	30/12/2022	30/12/2022
							F1908	Prolonged Surgery	A071204	No Pacing	30/12/2022	30/12/2022
					CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A071204	No Pacing	30/12/2022	30/12/2022
							F1908	Prolonged Surgery	A071204	No Pacing	30/12/2022	30/12/2022
					CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A071204	No Pacing	30/12/2022	30/12/2022
							F1908	Prolonged Surgery	A071204	No Pacing	30/12/2022	30/12/2022
					CT233	Surgical	E2403	No Clinical Signs, Symptoms or Conditions	A071204	No Pacing	30/12/2022	30/12/2022
							F1908	Prolonged Surgery	A071204	No Pacing	30/12/2022	30/12/2022
					CT2406	Implantable	E2403	No Clinical Signs, Symptoms or Conditions	A071204	No Pacing	30/12/2022	30/12/2022
							F1908	Prolonged Surgery	A071204	No Pacing	30/12/2022	30/12/2022
					CT2407	Active implantable	E2403	No Clinical Signs, Symptoms or Conditions	A071204	No Pacing	30/12/2022	30/12/2022
							F1908	Prolonged Surgery	A071204	No Pacing	30/12/2022	30/12/2022
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A071204	No Pacing	30/12/2022	30/12/2022
							F1908	Prolonged Surgery	A071204	No Pacing	30/12/2022	30/12/2022
					CT2727	Active implant devices/ancillaries	E2403	No Clinical Signs, Symptoms or Conditions	A071204	No Pacing	30/12/2022	30/12/2022
							F1908	Prolonged Surgery	A071204	No Pacing	30/12/2022	30/12/2022
					CT2750	Implantable cardiac electrical stimulators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A071204	No Pacing	30/12/2022	30/12/2022
							F1908	Prolonged Surgery	A071204	No Pacing	30/12/2022	30/12/2022
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A071204	No Pacing	30/12/2022	30/12/2022
							F1908	Prolonged Surgery	A071204	No Pacing	30/12/2022	30/12/2022
					CT2985	Active	E2403	No Clinical Signs, Symptoms or Conditions	A071204	No Pacing	30/12/2022	30/12/2022
							F1908	Prolonged Surgery	A071204	No Pacing	30/12/2022	30/12/2022
					CT3001	Implantable metal	E2403	No Clinical Signs, Symptoms or Conditions	A071204	No Pacing	30/12/2022	30/12/2022
							F1908	Prolonged Surgery	A071204	No Pacing	30/12/2022	30/12/2022
					CT3002	Implantable synthetic polymer	E2403	No Clinical Signs, Symptoms or Conditions	A071204	No Pacing	30/12/2022	30/12/2022
							F1908	Prolonged Surgery	A071204	No Pacing	30/12/2022	30/12/2022
					CT3006	Implantable inorganic material	E2403	No Clinical Signs, Symptoms or Conditions	A071204	No Pacing	30/12/2022	30/12/2022
							F1908	Prolonged Surgery	A071204	No Pacing	30/12/2022	30/12/2022
					CT314	Electrically powered	E2403	No Clinical Signs, Symptoms or Conditions	A071204	No Pacing	30/12/2022	30/12/2022
							F1908	Prolonged Surgery	A071204	No Pacing	30/12/2022	30/12/2022
					CT321	Long-term surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A071204	No Pacing	30/12/2022	30/12/2022
							F1908	Prolonged Surgery	A071204	No Pacing	30/12/2022	30/12/2022
					CT449	Generators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A071204	No Pacing	30/12/2022	30/12/2022
							F1908	Prolonged Surgery	A071204	No Pacing	30/12/2022	30/12/2022
					CT981	Single use	E2403	No Clinical Signs, Symptoms or Conditions	A071204	No Pacing	30/12/2022	30/12/2022
							F1908	Prolonged Surgery	A071204	No Pacing	30/12/2022	30/12/2022
					CT983	Surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A071204	No Pacing	30/12/2022	30/12/2022
							F1908	Prolonged Surgery	A071204	No Pacing	30/12/2022	30/12/2022
12060	2023	All other reportable incidents	35223	Endocardial/interventricular septal pacing lead	CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A0401	Break		
							F1901	Additional Surgery	A0401	Break		
12061	2023		47265	Dual-chamber implantable pacemaker, rate-responsive	CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	26/09/2022	26/09/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	26/09/2022	26/09/2022
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	26/09/2022	26/09/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	26/09/2022	26/09/2022
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	26/09/2022	26/09/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	26/09/2022	26/09/2022
12062	2023		47265	Dual-chamber implantable pacemaker, rate-responsive	CT141	Electrodes/Leads and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	17/01/2023	17/01/2023
							F26	No Health Consequences or Impact	A24	No Apparent Device Problem	17/01/2023	17/01/2023
					CT145	Cardiovascular devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	17/01/2023	17/01/2023
							F26	No Health Consequences or Impact	A24	No Apparent Device Problem	17/01/2023	17/01/2023
					CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	17/01/2023	17/01/2023
							F26	No Health Consequences or Impact	A24	No Apparent Device Problem	17/01/2023	17/01/2023
					CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	17/01/2023	17/01/2023
							F26	No Health Consequences or Impact	A24	No Apparent Device Problem	17/01/2023	17/01/2023
					CT233	Surgical	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	17/01/2023	17/01/2023
							F26	No Health Consequences or Impact	A24	No Apparent Device Problem	17/01/2023	17/01/2023
					CT2406	Implantable	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	17/01/2023	17/01/2023
							F26	No Health Consequences or Impact	A24	No Apparent Device Problem	17/01/2023	17/01/2023
					CT2407	Active implantable	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	17/01/2023	17/01/2023
							F26	No Health Consequences or Impact	A24	No Apparent Device Problem	17/01/2023	17/01/2023
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	17/01/2023	17/01/2023
							F26	No Health Consequences or Impact	A24	No Apparent Device Problem	17/01/2023	17/01/2023
					CT2727	Active implant devices/ancillaries	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	17/01/2023	17/01/2023
							F26	No Health Consequences or Impact	A24	No Apparent Device Problem	17/01/2023	17/01/2023
					CT2750	Implantable cardiac electrical stimulators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	17/01/2023	17/01/2023
							F26	No Health Consequences or Impact	A24	No Apparent Device Problem	17/01/2023	17/01/2023
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	17/01/2023	17/01/2023
							F26	No Health Consequences or Impact	A24	No Apparent Device Problem	17/01/2023	17/01/2023
					CT2985	Active	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	17/01/2023	17/01/2023
							F26	No Health Consequences or Impact	A24	No Apparent Device Problem	17/01/2023	17/01/2023
					CT3001	Implantable metal	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	17/01/2023	17/01/2023
							F26	No Health Consequences or Impact	A24	No Apparent Device Problem	17/01/2023	17/01/2023
					CT3002	Implantable synthetic polymer	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	17/01/2023	17/01/2023
							F26	No Health Consequences or Impact	A24	No Apparent Device Problem	17/01/2023	17/01/2023
					CT3006	Implantable inorganic material	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	17/01/2023	17/01/2023
							F26	No Health Consequences or Impact	A24	No Apparent Device Problem	17/01/2023	17/01/2023
					CT314	Electrically powered	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	17/01/2023	17/01/2023
							F26	No Health Consequences or Impact	A24	No Apparent Device Problem	17/01/2023	17/01/2023
					CT321	Long-term surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	17/01/2023	17/01/2023
							F26	No Health Consequences or Impact	A24	No Apparent Device Problem	17/01/2023	17/01/2023
					CT449	Generators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	17/01/2023	17/01/2023
							F26	No Health Consequences or Impact	A24	No Apparent Device Problem	17/01/2023	17/01/2023
					CT981	Single use	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	17/01/2023	17/01/2023
							F26	No Health Consequences or Impact	A24	No Apparent Device Problem	17/01/2023	17/01/2023
					CT983	Surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	17/01/2023	17/01/2023
							F26	No Health Consequences or Impact	A24	No Apparent Device Problem	17/01/2023	17/01/2023

12063	2023		47265	Dual-chamber implantable pacemaker, rate-responsive	CT141	Electrodes/Leads and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	11/01/2023	11/01/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	11/01/2023	11/01/2023
					CT145	Cardiovascular devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	11/01/2023	11/01/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	11/01/2023	11/01/2023
					CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	11/01/2023	11/01/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	11/01/2023	11/01/2023
					CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	11/01/2023	11/01/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	11/01/2023	11/01/2023
					CT233	Surgical	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	11/01/2023	11/01/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	11/01/2023	11/01/2023
					CT2406	Implantable	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	11/01/2023	11/01/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	11/01/2023	11/01/2023
					CT2407	Active implantable	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	11/01/2023	11/01/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	11/01/2023	11/01/2023
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	11/01/2023	11/01/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	11/01/2023	11/01/2023
					CT2727	Active implant devices/ancillaries	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	11/01/2023	11/01/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	11/01/2023	11/01/2023
					CT2750	Implantable cardiac electrical stimulators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	11/01/2023	11/01/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	11/01/2023	11/01/2023
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	11/01/2023	11/01/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	11/01/2023	11/01/2023
					CT2985	Active	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	11/01/2023	11/01/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	11/01/2023	11/01/2023
					CT3001	Implantable metal	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	11/01/2023	11/01/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	11/01/2023	11/01/2023
					CT3002	Implantable synthetic polymer	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	11/01/2023	11/01/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	11/01/2023	11/01/2023
					CT3006	Implantable inorganic material	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	11/01/2023	11/01/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	11/01/2023	11/01/2023
					CT314	Electrically powered	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	11/01/2023	11/01/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	11/01/2023	11/01/2023
					CT321	Long-term surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	11/01/2023	11/01/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	11/01/2023	11/01/2023
					CT449	Generators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	11/01/2023	11/01/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	11/01/2023	11/01/2023
					CT981	Single use	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	11/01/2023	11/01/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	11/01/2023	11/01/2023
					CT983	Surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	11/01/2023	11/01/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	11/01/2023	11/01/2023
12064	2023	All other reportable incidents	47263	Cardiac resynchronization therapy implantable pacemaker	CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A0908	Incorrect, Inadequate or Imprecise Result or Readings	23/12/2022	23/12/2022
							F08	Hospitalization or Prolonged Hospitalization	A0908	Incorrect, Inadequate or Imprecise Result or Readings	23/12/2022	23/12/2022
							F1905	Device Revision or Replacement	A0908	Incorrect, Inadequate or Imprecise Result or Readings	23/12/2022	23/12/2022
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A0908	Incorrect, Inadequate or Imprecise Result or Readings	23/12/2022	23/12/2022
							F08	Hospitalization or Prolonged Hospitalization	A0908	Incorrect, Inadequate or Imprecise Result or Readings	23/12/2022	23/12/2022
							F1905	Device Revision or Replacement	A0908	Incorrect, Inadequate or Imprecise Result or Readings	23/12/2022	23/12/2022
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A0908	Incorrect, Inadequate or Imprecise Result or Readings	23/12/2022	23/12/2022
							F08	Hospitalization or Prolonged Hospitalization	A0908	Incorrect, Inadequate or Imprecise Result or Readings	23/12/2022	23/12/2022
							F1905	Device Revision or Replacement	A0908	Incorrect, Inadequate or Imprecise Result or Readings	23/12/2022	23/12/2022
12065	2023		47265	Dual-chamber implantable pacemaker, rate-responsive	CT141	Electrodes/Leads and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	30/12/2022	30/12/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	30/12/2022	30/12/2022
					CT145	Cardiovascular devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	30/12/2022	30/12/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	30/12/2022	30/12/2022
					CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	30/12/2022	30/12/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	30/12/2022	30/12/2022
					CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	30/12/2022	30/12/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	30/12/2022	30/12/2022
					CT233	Surgical	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	30/12/2022	30/12/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	30/12/2022	30/12/2022
					CT2406	Implantable	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	30/12/2022	30/12/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	30/12/2022	30/12/2022
					CT2407	Active implantable	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	30/12/2022	30/12/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	30/12/2022	30/12/2022
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	30/12/2022	30/12/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	30/12/2022	30/12/2022
					CT2727	Active implant devices/ancillaries	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	30/12/2022	30/12/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	30/12/2022	30/12/2022
					CT2750	Implantable cardiac electrical stimulators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	30/12/2022	30/12/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	30/12/2022	30/12/2022
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	30/12/2022	30/12/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	30/12/2022	30/12/2022
					CT2985	Active	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	30/12/2022	30/12/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	30/12/2022	30/12/2022
					CT3001	Implantable metal	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	30/12/2022	30/12/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	30/12/2022	30/12/2022
					CT3002	Implantable synthetic polymer	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	30/12/2022	30/12/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	30/12/2022	30/12/2022
					CT3006	Implantable inorganic material	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	30/12/2022	30/12/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	30/12/2022	30/12/2022
					CT314	Electrically powered	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	30/12/2022	30/12/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	30/12/2022	30/12/2022

					CT321	Long-term surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	30/12/2022	30/12/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	30/12/2022	30/12/2022
					CT449	Generators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	30/12/2022	30/12/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	30/12/2022	30/12/2022
					CT981	Single use	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	30/12/2022	30/12/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	30/12/2022	30/12/2022
					CT983	Surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	30/12/2022	30/12/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	30/12/2022	30/12/2022
12066	2023		47265	Dual-chamber implantable pacemaker, rate-responsive	CT141	Electrodes/Leads and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	22/11/2022	22/11/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	22/11/2022	22/11/2022
					CT145	Cardiovascular devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	22/11/2022	22/11/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	22/11/2022	22/11/2022
					CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	22/11/2022	22/11/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	22/11/2022	22/11/2022
					CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	22/11/2022	22/11/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	22/11/2022	22/11/2022
					CT233	Surgical	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	22/11/2022	22/11/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	22/11/2022	22/11/2022
					CT2406	Implantable	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	22/11/2022	22/11/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	22/11/2022	22/11/2022
					CT2407	Active implantable	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	22/11/2022	22/11/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	22/11/2022	22/11/2022
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	22/11/2022	22/11/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	22/11/2022	22/11/2022
					CT2727	Active implant devices/ancillaries	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	22/11/2022	22/11/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	22/11/2022	22/11/2022
					CT2750	Implantable cardiac electrical stimulators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	22/11/2022	22/11/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	22/11/2022	22/11/2022
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	22/11/2022	22/11/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	22/11/2022	22/11/2022
					CT2985	Active	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	22/11/2022	22/11/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	22/11/2022	22/11/2022
					CT3001	Implantable metal	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	22/11/2022	22/11/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	22/11/2022	22/11/2022
					CT3002	Implantable synthetic polymer	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	22/11/2022	22/11/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	22/11/2022	22/11/2022
					CT3006	Implantable inorganic material	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	22/11/2022	22/11/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	22/11/2022	22/11/2022
					CT314	Electrically powered	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	22/11/2022	22/11/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	22/11/2022	22/11/2022
					CT321	Long-term surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	22/11/2022	22/11/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	22/11/2022	22/11/2022
					CT449	Generators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	22/11/2022	22/11/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	22/11/2022	22/11/2022
					CT981	Single use	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	22/11/2022	22/11/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	22/11/2022	22/11/2022
					CT983	Surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	22/11/2022	22/11/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	22/11/2022	22/11/2022
12067	2023		47265	Dual-chamber implantable pacemaker, rate-responsive	CT141	Electrodes/Leads and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	26/10/2022	26/10/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	26/10/2022	26/10/2022
					CT145	Cardiovascular devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	26/10/2022	26/10/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	26/10/2022	26/10/2022
					CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	26/10/2022	26/10/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	26/10/2022	26/10/2022
					CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	26/10/2022	26/10/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	26/10/2022	26/10/2022
					CT233	Surgical	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	26/10/2022	26/10/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	26/10/2022	26/10/2022
					CT2406	Implantable	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	26/10/2022	26/10/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	26/10/2022	26/10/2022
					CT2407	Active implantable	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	26/10/2022	26/10/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	26/10/2022	26/10/2022
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	26/10/2022	26/10/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	26/10/2022	26/10/2022
					CT2727	Active implant devices/ancillaries	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	26/10/2022	26/10/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	26/10/2022	26/10/2022
					CT2750	Implantable cardiac electrical stimulators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	26/10/2022	26/10/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	26/10/2022	26/10/2022
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	26/10/2022	26/10/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	26/10/2022	26/10/2022
					CT2985	Active	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	26/10/2022	26/10/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	26/10/2022	26/10/2022
					CT3001	Implantable metal	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	26/10/2022	26/10/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	26/10/2022	26/10/2022
					CT3002	Implantable synthetic polymer	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	26/10/2022	26/10/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	26/10/2022	26/10/2022
					CT3006	Implantable inorganic material	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	26/10/2022	26/10/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	26/10/2022	26/10/2022
					CT314	Electrically powered	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	26/10/2022	26/10/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	26/10/2022	26/10/2022
					CT321	Long-term surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	26/10/2022	26/10/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	26/10/2022	26/10/2022
					CT449	Generators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	26/10/2022	26/10/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	26/10/2022	26/10/2022
					CT981	Single use	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	26/10/2022	26/10/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	26/10/2022	26/10/2022
					CT983	Surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	26/10/2022	26/10/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	26/10/2022	26/10/2022

12068	2023		47265	Dual-chamber implantable pacemaker, rate-responsive	CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	05/09/2022	05/09/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	05/09/2022	05/09/2022
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	05/09/2022	05/09/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	05/09/2022	05/09/2022
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	05/09/2022	05/09/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	05/09/2022	05/09/2022
12069	2023		47265	Dual-chamber implantable pacemaker, rate-responsive	CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	29/11/2022	29/11/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	29/11/2022	29/11/2022
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	29/11/2022	29/11/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	29/11/2022	29/11/2022
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	29/11/2022	29/11/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	29/11/2022	29/11/2022
12070	2023	All other reportable incidents	47267	Single-chamber implantable pacemaker, rate-responsive	CT141	Electrodes/Leads and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT145	Cardiovascular devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT233	Surgical	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT2406	Implantable	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT2407	Active implantable	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT2727	Active implant devices/ancillaries	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT2750	Implantable cardiac electrical stimulators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT2985	Active	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT3001	Implantable metal	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT3002	Implantable synthetic polymer	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT3006	Implantable inorganic material	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT314	Electrically powered	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT321	Long-term surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT449	Generators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT981	Single use	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT983	Surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
12071	2023		47265	Dual-chamber implantable pacemaker, rate-responsive	CT141	Electrodes/Leads and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	06/01/2023	06/01/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	06/01/2023	06/01/2023
					CT145	Cardiovascular devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	06/01/2023	06/01/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	06/01/2023	06/01/2023
					CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	06/01/2023	06/01/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	06/01/2023	06/01/2023
					CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	06/01/2023	06/01/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	06/01/2023	06/01/2023
					CT233	Surgical	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	06/01/2023	06/01/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	06/01/2023	06/01/2023
					CT2406	Implantable	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	06/01/2023	06/01/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	06/01/2023	06/01/2023
					CT2407	Active implantable	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	06/01/2023	06/01/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	06/01/2023	06/01/2023
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	06/01/2023	06/01/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	06/01/2023	06/01/2023
					CT2727	Active implant devices/ancillaries	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	06/01/2023	06/01/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	06/01/2023	06/01/2023
					CT2750	Implantable cardiac electrical stimulators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	06/01/2023	06/01/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	06/01/2023	06/01/2023
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	06/01/2023	06/01/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	06/01/2023	06/01/2023
					CT2985	Active	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	06/01/2023	06/01/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	06/01/2023	06/01/2023
					CT3001	Implantable metal	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	06/01/2023	06/01/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	06/01/2023	06/01/2023
					CT3002	Implantable synthetic polymer	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	06/01/2023	06/01/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	06/01/2023	06/01/2023
					CT3006	Implantable inorganic material	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	06/01/2023	06/01/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	06/01/2023	06/01/2023
					CT314	Electrically powered	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	06/01/2023	06/01/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	06/01/2023	06/01/2023
					CT321	Long-term surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	06/01/2023	06/01/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	06/01/2023	06/01/2023
					CT449	Generators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	06/01/2023	06/01/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	06/01/2023	06/01/2023

					CT981	Single use	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	06/01/2023	06/01/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	06/01/2023	06/01/2023
					CT983	Surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	06/01/2023	06/01/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	06/01/2023	06/01/2023
12072	2023	All other reportable incidents	47265	Dual-chamber implantable pacemaker, rate-responsive	CT141	Electrodes/Leads and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	17/01/2023	17/01/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	17/01/2023	17/01/2023
					CT145	Cardiovascular devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	17/01/2023	17/01/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	17/01/2023	17/01/2023
					CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	17/01/2023	17/01/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	17/01/2023	17/01/2023
					CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	17/01/2023	17/01/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	17/01/2023	17/01/2023
					CT233	Surgical	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	17/01/2023	17/01/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	17/01/2023	17/01/2023
					CT2406	Implantable	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	17/01/2023	17/01/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	17/01/2023	17/01/2023
					CT2407	Active implantable	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	17/01/2023	17/01/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	17/01/2023	17/01/2023
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	17/01/2023	17/01/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	17/01/2023	17/01/2023
					CT2727	Active implant devices/ancillaries	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	17/01/2023	17/01/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	17/01/2023	17/01/2023
					CT2750	Implantable cardiac electrical stimulators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	17/01/2023	17/01/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	17/01/2023	17/01/2023
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	17/01/2023	17/01/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	17/01/2023	17/01/2023
					CT2985	Active	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	17/01/2023	17/01/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	17/01/2023	17/01/2023
					CT3001	Implantable metal	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	17/01/2023	17/01/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	17/01/2023	17/01/2023
					CT3002	Implantable synthetic polymer	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	17/01/2023	17/01/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	17/01/2023	17/01/2023
					CT3006	Implantable inorganic material	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	17/01/2023	17/01/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	17/01/2023	17/01/2023
					CT314	Electrically powered	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	17/01/2023	17/01/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	17/01/2023	17/01/2023
					CT321	Long-term surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	17/01/2023	17/01/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	17/01/2023	17/01/2023
					CT449	Generators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	17/01/2023	17/01/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	17/01/2023	17/01/2023
					CT981	Single use	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	17/01/2023	17/01/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	17/01/2023	17/01/2023
					CT983	Surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	17/01/2023	17/01/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	17/01/2023	17/01/2023
12073	2023	All other reportable incidents	47265	Dual-chamber implantable pacemaker, rate-responsive	CT1685	Cardiac pacemakers	E011903	Syncope/Fainting	A0712	Pacing Problem	20/01/2023	20/01/2023
							F08	Hospitalization or Prolonged Hospitalization	A0712	Pacing Problem	20/01/2023	20/01/2023
							F1203	Life Threatening Illness or Injury	A0712	Pacing Problem	20/01/2023	20/01/2023
							F1905	Device Revision or Replacement	A0712	Pacing Problem	20/01/2023	20/01/2023
					CT257	Implantable cardiac pacemakers	E011903	Syncope/Fainting	A0712	Pacing Problem	20/01/2023	20/01/2023
							F08	Hospitalization or Prolonged Hospitalization	A0712	Pacing Problem	20/01/2023	20/01/2023
							F1203	Life Threatening Illness or Injury	A0712	Pacing Problem	20/01/2023	20/01/2023
							F1905	Device Revision or Replacement	A0712	Pacing Problem	20/01/2023	20/01/2023
					CT2866	Implantable cardiac electrical stimulators	E011903	Syncope/Fainting	A0712	Pacing Problem	20/01/2023	20/01/2023
							F08	Hospitalization or Prolonged Hospitalization	A0712	Pacing Problem	20/01/2023	20/01/2023
							F1203	Life Threatening Illness or Injury	A0712	Pacing Problem	20/01/2023	20/01/2023
							F1905	Device Revision or Replacement	A0712	Pacing Problem	20/01/2023	20/01/2023
12074	2023		47265	Dual-chamber implantable pacemaker, rate-responsive	CT141	Electrodes/Leads and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	09/11/2022	09/11/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	09/11/2022	09/11/2022
					CT145	Cardiovascular devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	09/11/2022	09/11/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	09/11/2022	09/11/2022
					CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	09/11/2022	09/11/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	09/11/2022	09/11/2022
					CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	09/11/2022	09/11/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	09/11/2022	09/11/2022
					CT233	Surgical	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	09/11/2022	09/11/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	09/11/2022	09/11/2022
					CT2406	Implantable	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	09/11/2022	09/11/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	09/11/2022	09/11/2022
					CT2407	Active implantable	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	09/11/2022	09/11/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	09/11/2022	09/11/2022
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	09/11/2022	09/11/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	09/11/2022	09/11/2022
					CT2727	Active implant devices/ancillaries	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	09/11/2022	09/11/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	09/11/2022	09/11/2022
					CT2750	Implantable cardiac electrical stimulators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	09/11/2022	09/11/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	09/11/2022	09/11/2022
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	09/11/2022	09/11/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	09/11/2022	09/11/2022
					CT2985	Active	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	09/11/2022	09/11/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	09/11/2022	09/11/2022
					CT3001	Implantable metal	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	09/11/2022	09/11/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	09/11/2022	09/11/2022
					CT3002	Implantable synthetic polymer	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	09/11/2022	09/11/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	09/11/2022	09/11/2022
					CT3006	Implantable inorganic material	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	09/11/2022	09/11/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	09/11/2022	09/11/2022
					CT314	Electrically powered	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	09/11/2022	09/11/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	09/11/2022	09/11/2022

					CT321	Long-term surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	09/11/2022	09/11/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	09/11/2022	09/11/2022
					CT449	Generators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	09/11/2022	09/11/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	09/11/2022	09/11/2022
					CT981	Single use	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	09/11/2022	09/11/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	09/11/2022	09/11/2022
					CT983	Surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	09/11/2022	09/11/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	09/11/2022	09/11/2022
12075	2023		47265	Dual-chamber implantable pacemaker, rate-responsive	CT141	Electrodes/Leads and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	28/09/2022	28/09/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	28/09/2022	28/09/2022
					CT145	Cardiovascular devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	28/09/2022	28/09/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	28/09/2022	28/09/2022
					CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	28/09/2022	28/09/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	28/09/2022	28/09/2022
					CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	28/09/2022	28/09/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	28/09/2022	28/09/2022
					CT233	Surgical	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	28/09/2022	28/09/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	28/09/2022	28/09/2022
					CT2406	Implantable	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	28/09/2022	28/09/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	28/09/2022	28/09/2022
					CT2407	Active implantable	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	28/09/2022	28/09/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	28/09/2022	28/09/2022
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	28/09/2022	28/09/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	28/09/2022	28/09/2022
					CT2727	Active implant devices/ancillaries	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	28/09/2022	28/09/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	28/09/2022	28/09/2022
					CT2750	Implantable cardiac electrical stimulators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	28/09/2022	28/09/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	28/09/2022	28/09/2022
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	28/09/2022	28/09/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	28/09/2022	28/09/2022
					CT2985	Active	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	28/09/2022	28/09/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	28/09/2022	28/09/2022
					CT3001	Implantable metal	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	28/09/2022	28/09/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	28/09/2022	28/09/2022
					CT3002	Implantable synthetic polymer	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	28/09/2022	28/09/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	28/09/2022	28/09/2022
					CT3006	Implantable inorganic material	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	28/09/2022	28/09/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	28/09/2022	28/09/2022
					CT314	Electrically powered	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	28/09/2022	28/09/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	28/09/2022	28/09/2022
					CT321	Long-term surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	28/09/2022	28/09/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	28/09/2022	28/09/2022
					CT449	Generators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	28/09/2022	28/09/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	28/09/2022	28/09/2022
					CT981	Single use	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	28/09/2022	28/09/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	28/09/2022	28/09/2022
					CT983	Surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	28/09/2022	28/09/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	28/09/2022	28/09/2022
12076	2023		47265	Dual-chamber implantable pacemaker, rate-responsive	CT141	Electrodes/Leads and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	30/11/2022	30/11/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	30/11/2022	30/11/2022
					CT145	Cardiovascular devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	30/11/2022	30/11/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	30/11/2022	30/11/2022
					CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	30/11/2022	30/11/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	30/11/2022	30/11/2022
					CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	30/11/2022	30/11/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	30/11/2022	30/11/2022
					CT233	Surgical	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	30/11/2022	30/11/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	30/11/2022	30/11/2022
					CT2406	Implantable	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	30/11/2022	30/11/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	30/11/2022	30/11/2022
					CT2407	Active implantable	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	30/11/2022	30/11/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	30/11/2022	30/11/2022
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	30/11/2022	30/11/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	30/11/2022	30/11/2022
					CT2727	Active implant devices/ancillaries	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	30/11/2022	30/11/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	30/11/2022	30/11/2022
					CT2750	Implantable cardiac electrical stimulators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	30/11/2022	30/11/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	30/11/2022	30/11/2022
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	30/11/2022	30/11/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	30/11/2022	30/11/2022
					CT2985	Active	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	30/11/2022	30/11/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	30/11/2022	30/11/2022
					CT3001	Implantable metal	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	30/11/2022	30/11/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	30/11/2022	30/11/2022
					CT3002	Implantable synthetic polymer	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	30/11/2022	30/11/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	30/11/2022	30/11/2022
					CT3006	Implantable inorganic material	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	30/11/2022	30/11/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	30/11/2022	30/11/2022
					CT314	Electrically powered	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	30/11/2022	30/11/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	30/11/2022	30/11/2022
					CT321	Long-term surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	30/11/2022	30/11/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	30/11/2022	30/11/2022
					CT449	Generators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	30/11/2022	30/11/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	30/11/2022	30/11/2022
					CT981	Single use	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	30/11/2022	30/11/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	30/11/2022	30/11/2022
					CT983	Surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	30/11/2022	30/11/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	30/11/2022	30/11/2022

12077	2023		47265	Dual-chamber implantable pacemaker, rate-responsive	CT141	Electrodes/Leads and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	10/01/2023	10/01/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	10/01/2023	10/01/2023
					CT145	Cardiovascular devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	10/01/2023	10/01/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	10/01/2023	10/01/2023
					CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	10/01/2023	10/01/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	10/01/2023	10/01/2023
					CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	10/01/2023	10/01/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	10/01/2023	10/01/2023
					CT233	Surgical	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	10/01/2023	10/01/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	10/01/2023	10/01/2023
					CT2406	Implantable	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	10/01/2023	10/01/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	10/01/2023	10/01/2023
					CT2407	Active implantable	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	10/01/2023	10/01/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	10/01/2023	10/01/2023
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	10/01/2023	10/01/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	10/01/2023	10/01/2023
					CT2727	Active implant devices/ancillaries	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	10/01/2023	10/01/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	10/01/2023	10/01/2023
					CT2750	Implantable cardiac electrical stimulators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	10/01/2023	10/01/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	10/01/2023	10/01/2023
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	10/01/2023	10/01/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	10/01/2023	10/01/2023
					CT2985	Active	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	10/01/2023	10/01/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	10/01/2023	10/01/2023
					CT3001	Implantable metal	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	10/01/2023	10/01/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	10/01/2023	10/01/2023
					CT3002	Implantable synthetic polymer	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	10/01/2023	10/01/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	10/01/2023	10/01/2023
					CT3006	Implantable inorganic material	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	10/01/2023	10/01/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	10/01/2023	10/01/2023
					CT314	Electrically powered	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	10/01/2023	10/01/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	10/01/2023	10/01/2023
					CT321	Long-term surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	10/01/2023	10/01/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	10/01/2023	10/01/2023
					CT449	Generators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	10/01/2023	10/01/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	10/01/2023	10/01/2023
					CT981	Single use	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	10/01/2023	10/01/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	10/01/2023	10/01/2023
					CT983	Surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	10/01/2023	10/01/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	10/01/2023	10/01/2023
12078	2023		47265	Dual-chamber implantable pacemaker, rate-responsive	CT141	Electrodes/Leads and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	05/12/2022	05/12/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	05/12/2022	05/12/2022
					CT145	Cardiovascular devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	05/12/2022	05/12/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	05/12/2022	05/12/2022
					CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	05/12/2022	05/12/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	05/12/2022	05/12/2022
					CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	05/12/2022	05/12/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	05/12/2022	05/12/2022
					CT233	Surgical	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	05/12/2022	05/12/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	05/12/2022	05/12/2022
					CT2406	Implantable	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	05/12/2022	05/12/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	05/12/2022	05/12/2022
					CT2407	Active implantable	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	05/12/2022	05/12/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	05/12/2022	05/12/2022
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	05/12/2022	05/12/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	05/12/2022	05/12/2022
					CT2727	Active implant devices/ancillaries	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	05/12/2022	05/12/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	05/12/2022	05/12/2022
					CT2750	Implantable cardiac electrical stimulators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	05/12/2022	05/12/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	05/12/2022	05/12/2022
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	05/12/2022	05/12/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	05/12/2022	05/12/2022
					CT2985	Active	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	05/12/2022	05/12/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	05/12/2022	05/12/2022
					CT3001	Implantable metal	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	05/12/2022	05/12/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	05/12/2022	05/12/2022
					CT3002	Implantable synthetic polymer	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	05/12/2022	05/12/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	05/12/2022	05/12/2022
					CT3006	Implantable inorganic material	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	05/12/2022	05/12/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	05/12/2022	05/12/2022
					CT314	Electrically powered	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	05/12/2022	05/12/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	05/12/2022	05/12/2022
					CT321	Long-term surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	05/12/2022	05/12/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	05/12/2022	05/12/2022
					CT449	Generators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	05/12/2022	05/12/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	05/12/2022	05/12/2022
					CT981	Single use	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	05/12/2022	05/12/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	05/12/2022	05/12/2022
					CT983	Surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	05/12/2022	05/12/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	05/12/2022	05/12/2022
12079	2023		47265	Dual-chamber implantable pacemaker, rate-responsive	CT141	Electrodes/Leads and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	05/12/2022	05/12/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	05/12/2022	05/12/2022
					CT145	Cardiovascular devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	05/12/2022	05/12/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	05/12/2022	05/12/2022
					CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	05/12/2022	05/12/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	05/12/2022	05/12/2022
					CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	05/12/2022	05/12/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	05/12/2022	05/12/2022

					CT233	Surgical	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	05/12/2022	05/12/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	05/12/2022	05/12/2022
					CT2406	Implantable	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	05/12/2022	05/12/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	05/12/2022	05/12/2022
					CT2407	Active implantable	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	05/12/2022	05/12/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	05/12/2022	05/12/2022
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	05/12/2022	05/12/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	05/12/2022	05/12/2022
					CT2727	Active implant devices/ancillaries	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	05/12/2022	05/12/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	05/12/2022	05/12/2022
					CT2750	Implantable cardiac electrical stimulators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	05/12/2022	05/12/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	05/12/2022	05/12/2022
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	05/12/2022	05/12/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	05/12/2022	05/12/2022
					CT2985	Active	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	05/12/2022	05/12/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	05/12/2022	05/12/2022
					CT3001	Implantable metal	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	05/12/2022	05/12/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	05/12/2022	05/12/2022
					CT3002	Implantable synthetic polymer	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	05/12/2022	05/12/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	05/12/2022	05/12/2022
					CT3006	Implantable inorganic material	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	05/12/2022	05/12/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	05/12/2022	05/12/2022
					CT314	Electrically powered	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	05/12/2022	05/12/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	05/12/2022	05/12/2022
					CT321	Long-term surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	05/12/2022	05/12/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	05/12/2022	05/12/2022
					CT449	Generators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	05/12/2022	05/12/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	05/12/2022	05/12/2022
					CT981	Single use	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	05/12/2022	05/12/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	05/12/2022	05/12/2022
					CT983	Surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	05/12/2022	05/12/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	05/12/2022	05/12/2022
12080	2023	All other reportable incidents	17882	Physiologic-monitoring defibrillation system	CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A0719	Unexpected Shutdown		
							F27	Problem Identified During Non-Clinical Procedure	A0719	Unexpected Shutdown		
12081	2023		47265	Dual-chamber implantable pacemaker, rate-responsive	CT141	Electrodes/Leads and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	05/12/2022	05/12/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	05/12/2022	05/12/2022
					CT145	Cardiovascular devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	05/12/2022	05/12/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	05/12/2022	05/12/2022
					CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	05/12/2022	05/12/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	05/12/2022	05/12/2022
					CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	05/12/2022	05/12/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	05/12/2022	05/12/2022
					CT233	Surgical	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	05/12/2022	05/12/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	05/12/2022	05/12/2022
					CT2406	Implantable	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	05/12/2022	05/12/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	05/12/2022	05/12/2022
					CT2407	Active implantable	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	05/12/2022	05/12/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	05/12/2022	05/12/2022
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	05/12/2022	05/12/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	05/12/2022	05/12/2022
					CT2727	Active implant devices/ancillaries	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	05/12/2022	05/12/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	05/12/2022	05/12/2022
					CT2750	Implantable cardiac electrical stimulators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	05/12/2022	05/12/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	05/12/2022	05/12/2022
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	05/12/2022	05/12/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	05/12/2022	05/12/2022
					CT2985	Active	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	05/12/2022	05/12/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	05/12/2022	05/12/2022
					CT3001	Implantable metal	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	05/12/2022	05/12/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	05/12/2022	05/12/2022
					CT3002	Implantable synthetic polymer	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	05/12/2022	05/12/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	05/12/2022	05/12/2022
					CT3006	Implantable inorganic material	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	05/12/2022	05/12/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	05/12/2022	05/12/2022
					CT314	Electrically powered	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	05/12/2022	05/12/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	05/12/2022	05/12/2022
					CT321	Long-term surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	05/12/2022	05/12/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	05/12/2022	05/12/2022
					CT449	Generators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	05/12/2022	05/12/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	05/12/2022	05/12/2022
					CT981	Single use	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	05/12/2022	05/12/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	05/12/2022	05/12/2022
					CT983	Surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	05/12/2022	05/12/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	05/12/2022	05/12/2022
12082	2023	All other reportable incidents	47265	Dual-chamber implantable pacemaker, rate-responsive	CT141	Electrodes/Leads and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	29/11/2022	29/11/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	29/11/2022	29/11/2022
					CT145	Cardiovascular devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	29/11/2022	29/11/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	29/11/2022	29/11/2022
					CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	29/11/2022	29/11/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	29/11/2022	29/11/2022
					CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	29/11/2022	29/11/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	29/11/2022	29/11/2022
					CT233	Surgical	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	29/11/2022	29/11/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	29/11/2022	29/11/2022
					CT2406	Implantable	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	29/11/2022	29/11/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	29/11/2022	29/11/2022
					CT2407	Active implantable	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	29/11/2022	29/11/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	29/11/2022	29/11/2022

					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	29/11/2022	29/11/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	29/11/2022	29/11/2022
					CT2727	Active implant devices/ancillaries	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	29/11/2022	29/11/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	29/11/2022	29/11/2022
					CT2750	Implantable cardiac electrical stimulators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	29/11/2022	29/11/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	29/11/2022	29/11/2022
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	29/11/2022	29/11/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	29/11/2022	29/11/2022
					CT2985	Active	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	29/11/2022	29/11/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	29/11/2022	29/11/2022
					CT3001	Implantable metal	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	29/11/2022	29/11/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	29/11/2022	29/11/2022
					CT3002	Implantable synthetic polymer	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	29/11/2022	29/11/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	29/11/2022	29/11/2022
					CT3006	Implantable inorganic material	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	29/11/2022	29/11/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	29/11/2022	29/11/2022
					CT314	Electrically powered	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	29/11/2022	29/11/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	29/11/2022	29/11/2022
					CT321	Long-term surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	29/11/2022	29/11/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	29/11/2022	29/11/2022
					CT449	Generators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	29/11/2022	29/11/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	29/11/2022	29/11/2022
					CT981	Single use	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	29/11/2022	29/11/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	29/11/2022	29/11/2022
					CT983	Surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	29/11/2022	29/11/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	29/11/2022	29/11/2022
12083	2023		47265	Dual-chamber implantable pacemaker, rate-responsive	CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	13/12/2022	13/12/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	13/12/2022	13/12/2022
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	13/12/2022	13/12/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	13/12/2022	13/12/2022
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	13/12/2022	13/12/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	13/12/2022	13/12/2022
12084	2023		47265	Dual-chamber implantable pacemaker, rate-responsive	CT141	Electrodes/Leads and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem		
							F1901	Additional Surgery	A24	No Apparent Device Problem		
					CT145	Cardiovascular devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem		
							F1901	Additional Surgery	A24	No Apparent Device Problem		
					CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem		
							F1901	Additional Surgery	A24	No Apparent Device Problem		
					CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem		
							F1901	Additional Surgery	A24	No Apparent Device Problem		
					CT233	Surgical	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem		
							F1901	Additional Surgery	A24	No Apparent Device Problem		
					CT2406	Implantable	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem		
							F1901	Additional Surgery	A24	No Apparent Device Problem		
					CT2407	Active implantable	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem		
							F1901	Additional Surgery	A24	No Apparent Device Problem		
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem		
							F1901	Additional Surgery	A24	No Apparent Device Problem		
					CT2727	Active implant devices/ancillaries	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem		
							F1901	Additional Surgery	A24	No Apparent Device Problem		
					CT2750	Implantable cardiac electrical stimulators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem		
							F1901	Additional Surgery	A24	No Apparent Device Problem		
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem		
							F1901	Additional Surgery	A24	No Apparent Device Problem		
					CT2985	Active	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem		
							F1901	Additional Surgery	A24	No Apparent Device Problem		
					CT3001	Implantable metal	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem		
							F1901	Additional Surgery	A24	No Apparent Device Problem		
					CT3002	Implantable synthetic polymer	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem		
							F1901	Additional Surgery	A24	No Apparent Device Problem		
					CT3006	Implantable inorganic material	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem		
							F1901	Additional Surgery	A24	No Apparent Device Problem		
					CT314	Electrically powered	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem		
							F1901	Additional Surgery	A24	No Apparent Device Problem		
					CT321	Long-term surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem		
							F1901	Additional Surgery	A24	No Apparent Device Problem		
					CT449	Generators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem		
							F1901	Additional Surgery	A24	No Apparent Device Problem		
					CT981	Single use	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem		
							F1901	Additional Surgery	A24	No Apparent Device Problem		
					CT983	Surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem		
							F1901	Additional Surgery	A24	No Apparent Device Problem		
12085	2023		47265	Dual-chamber implantable pacemaker, rate-responsive	CT141	Electrodes/Leads and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	12/11/2022	12/11/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	12/11/2022	12/11/2022
					CT145	Cardiovascular devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	12/11/2022	12/11/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	12/11/2022	12/11/2022
					CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	12/11/2022	12/11/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	12/11/2022	12/11/2022
					CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	12/11/2022	12/11/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	12/11/2022	12/11/2022
					CT233	Surgical	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	12/11/2022	12/11/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	12/11/2022	12/11/2022
					CT2406	Implantable	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	12/11/2022	12/11/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	12/11/2022	12/11/2022
					CT2407	Active implantable	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	12/11/2022	12/11/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	12/11/2022	12/11/2022
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	12/11/2022	12/11/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	12/11/2022	12/11/2022

					CT2727	Active implant devices/ancillaries	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	12/11/2022	12/11/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	12/11/2022	12/11/2022
					CT2750	Implantable cardiac electrical stimulators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	12/11/2022	12/11/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	12/11/2022	12/11/2022
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	12/11/2022	12/11/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	12/11/2022	12/11/2022
					CT2985	Active	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	12/11/2022	12/11/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	12/11/2022	12/11/2022
					CT3001	Implantable metal	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	12/11/2022	12/11/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	12/11/2022	12/11/2022
					CT3002	Implantable synthetic polymer	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	12/11/2022	12/11/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	12/11/2022	12/11/2022
					CT3006	Implantable inorganic material	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	12/11/2022	12/11/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	12/11/2022	12/11/2022
					CT314	Electrically powered	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	12/11/2022	12/11/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	12/11/2022	12/11/2022
					CT321	Long-term surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	12/11/2022	12/11/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	12/11/2022	12/11/2022
					CT449	Generators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	12/11/2022	12/11/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	12/11/2022	12/11/2022
					CT981	Single use	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	12/11/2022	12/11/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	12/11/2022	12/11/2022
					CT983	Surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	12/11/2022	12/11/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	12/11/2022	12/11/2022
12086	2023		47265	Dual-chamber implantable pacemaker, rate-responsive	CT141	Electrodes/Leads and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	10/01/2023	10/01/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	10/01/2023	10/01/2023
					CT145	Cardiovascular devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	10/01/2023	10/01/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	10/01/2023	10/01/2023
					CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	10/01/2023	10/01/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	10/01/2023	10/01/2023
					CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	10/01/2023	10/01/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	10/01/2023	10/01/2023
					CT233	Surgical	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	10/01/2023	10/01/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	10/01/2023	10/01/2023
					CT2406	Implantable	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	10/01/2023	10/01/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	10/01/2023	10/01/2023
					CT2407	Active implantable	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	10/01/2023	10/01/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	10/01/2023	10/01/2023
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	10/01/2023	10/01/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	10/01/2023	10/01/2023
					CT2727	Active implant devices/ancillaries	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	10/01/2023	10/01/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	10/01/2023	10/01/2023
					CT2750	Implantable cardiac electrical stimulators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	10/01/2023	10/01/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	10/01/2023	10/01/2023
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	10/01/2023	10/01/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	10/01/2023	10/01/2023
					CT2985	Active	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	10/01/2023	10/01/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	10/01/2023	10/01/2023
					CT3001	Implantable metal	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	10/01/2023	10/01/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	10/01/2023	10/01/2023
					CT3002	Implantable synthetic polymer	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	10/01/2023	10/01/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	10/01/2023	10/01/2023
					CT3006	Implantable inorganic material	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	10/01/2023	10/01/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	10/01/2023	10/01/2023
					CT314	Electrically powered	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	10/01/2023	10/01/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	10/01/2023	10/01/2023
					CT321	Long-term surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	10/01/2023	10/01/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	10/01/2023	10/01/2023
					CT449	Generators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	10/01/2023	10/01/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	10/01/2023	10/01/2023
					CT981	Single use	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	10/01/2023	10/01/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	10/01/2023	10/01/2023
					CT983	Surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	10/01/2023	10/01/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	10/01/2023	10/01/2023
12087	2023	All other reportable incidents	47265	Dual-chamber implantable pacemaker, rate-responsive	CT141	Electrodes/Leads and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	13/01/2023	13/01/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	13/01/2023	13/01/2023
					CT145	Cardiovascular devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	13/01/2023	13/01/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	13/01/2023	13/01/2023
					CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	13/01/2023	13/01/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	13/01/2023	13/01/2023
					CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	13/01/2023	13/01/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	13/01/2023	13/01/2023
					CT233	Surgical	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	13/01/2023	13/01/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	13/01/2023	13/01/2023
					CT2406	Implantable	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	13/01/2023	13/01/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	13/01/2023	13/01/2023
					CT2407	Active implantable	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	13/01/2023	13/01/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	13/01/2023	13/01/2023
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	13/01/2023	13/01/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	13/01/2023	13/01/2023
					CT2727	Active implant devices/ancillaries	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	13/01/2023	13/01/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	13/01/2023	13/01/2023
					CT2750	Implantable cardiac electrical stimulators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	13/01/2023	13/01/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	13/01/2023	13/01/2023
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	13/01/2023	13/01/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	13/01/2023	13/01/2023
					CT2985	Active	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	13/01/2023	13/01/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	13/01/2023	13/01/2023

					CT3001	Implantable metal	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	13/01/2023	13/01/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	13/01/2023	13/01/2023
					CT3002	Implantable synthetic polymer	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	13/01/2023	13/01/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	13/01/2023	13/01/2023
					CT3006	Implantable inorganic material	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	13/01/2023	13/01/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	13/01/2023	13/01/2023
					CT314	Electrically powered	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	13/01/2023	13/01/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	13/01/2023	13/01/2023
					CT321	Long-term surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	13/01/2023	13/01/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	13/01/2023	13/01/2023
					CT449	Generators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	13/01/2023	13/01/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	13/01/2023	13/01/2023
					CT981	Single use	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	13/01/2023	13/01/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	13/01/2023	13/01/2023
					CT983	Surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	13/01/2023	13/01/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	13/01/2023	13/01/2023
12088	2023	All other reportable incidents	37265	Dual-chamber implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	08/11/2022	08/11/2022
							F1901	Additional Surgery	A070504	Premature Discharge of Battery	08/11/2022	08/11/2022
12089	2023		47265	Dual-chamber implantable pacemaker, rate-responsive	CT141	Electrodes/Leads and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	11/01/2023	11/01/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	11/01/2023	11/01/2023
					CT145	Cardiovascular devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	11/01/2023	11/01/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	11/01/2023	11/01/2023
					CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	11/01/2023	11/01/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	11/01/2023	11/01/2023
					CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	11/01/2023	11/01/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	11/01/2023	11/01/2023
					CT233	Surgical	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	11/01/2023	11/01/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	11/01/2023	11/01/2023
					CT2406	Implantable	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	11/01/2023	11/01/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	11/01/2023	11/01/2023
					CT2407	Active implantable	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	11/01/2023	11/01/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	11/01/2023	11/01/2023
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	11/01/2023	11/01/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	11/01/2023	11/01/2023
					CT2727	Active implant devices/ancillaries	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	11/01/2023	11/01/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	11/01/2023	11/01/2023
					CT2750	Implantable cardiac electrical stimulators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	11/01/2023	11/01/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	11/01/2023	11/01/2023
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	11/01/2023	11/01/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	11/01/2023	11/01/2023
					CT2985	Active	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	11/01/2023	11/01/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	11/01/2023	11/01/2023
					CT3001	Implantable metal	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	11/01/2023	11/01/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	11/01/2023	11/01/2023
					CT3002	Implantable synthetic polymer	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	11/01/2023	11/01/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	11/01/2023	11/01/2023
					CT3006	Implantable inorganic material	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	11/01/2023	11/01/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	11/01/2023	11/01/2023
					CT314	Electrically powered	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	11/01/2023	11/01/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	11/01/2023	11/01/2023
					CT321	Long-term surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	11/01/2023	11/01/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	11/01/2023	11/01/2023
					CT449	Generators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	11/01/2023	11/01/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	11/01/2023	11/01/2023
					CT981	Single use	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	11/01/2023	11/01/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	11/01/2023	11/01/2023
					CT983	Surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	11/01/2023	11/01/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	11/01/2023	11/01/2023
12090	2023		47265	Dual-chamber implantable pacemaker, rate-responsive	CT141	Electrodes/Leads and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	07/01/2023	07/01/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	07/01/2023	07/01/2023
					CT145	Cardiovascular devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	07/01/2023	07/01/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	07/01/2023	07/01/2023
					CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	07/01/2023	07/01/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	07/01/2023	07/01/2023
					CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	07/01/2023	07/01/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	07/01/2023	07/01/2023
					CT233	Surgical	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	07/01/2023	07/01/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	07/01/2023	07/01/2023
					CT2406	Implantable	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	07/01/2023	07/01/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	07/01/2023	07/01/2023
					CT2407	Active implantable	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	07/01/2023	07/01/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	07/01/2023	07/01/2023
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	07/01/2023	07/01/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	07/01/2023	07/01/2023
					CT2727	Active implant devices/ancillaries	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	07/01/2023	07/01/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	07/01/2023	07/01/2023
					CT2750	Implantable cardiac electrical stimulators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	07/01/2023	07/01/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	07/01/2023	07/01/2023
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	07/01/2023	07/01/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	07/01/2023	07/01/2023
					CT2985	Active	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	07/01/2023	07/01/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	07/01/2023	07/01/2023
					CT3001	Implantable metal	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	07/01/2023	07/01/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	07/01/2023	07/01/2023
					CT3002	Implantable synthetic polymer	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	07/01/2023	07/01/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	07/01/2023	07/01/2023
					CT3006	Implantable inorganic material	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	07/01/2023	07/01/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	07/01/2023	07/01/2023

					CT314	Electrically powered	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	07/01/2023	07/01/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	07/01/2023	07/01/2023
					CT321	Long-term surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	07/01/2023	07/01/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	07/01/2023	07/01/2023
					CT449	Generators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	07/01/2023	07/01/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	07/01/2023	07/01/2023
					CT981	Single use	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	07/01/2023	07/01/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	07/01/2023	07/01/2023
					CT983	Surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	07/01/2023	07/01/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	07/01/2023	07/01/2023
12091	2023		47265	Dual-chamber implantable pacemaker, rate-responsive	CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	23/12/2022	23/12/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	23/12/2022	23/12/2022
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	23/12/2022	23/12/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	23/12/2022	23/12/2022
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	23/12/2022	23/12/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	23/12/2022	23/12/2022
12092	2023		47265	Dual-chamber implantable pacemaker, rate-responsive	CT141	Electrodes/Leads and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	10/12/2022	10/12/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	10/12/2022	10/12/2022
					CT145	Cardiovascular devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	10/12/2022	10/12/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	10/12/2022	10/12/2022
					CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	10/12/2022	10/12/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	10/12/2022	10/12/2022
					CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	10/12/2022	10/12/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	10/12/2022	10/12/2022
					CT233	Surgical	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	10/12/2022	10/12/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	10/12/2022	10/12/2022
					CT2406	Implantable	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	10/12/2022	10/12/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	10/12/2022	10/12/2022
					CT2407	Active implantable	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	10/12/2022	10/12/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	10/12/2022	10/12/2022
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	10/12/2022	10/12/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	10/12/2022	10/12/2022
					CT2727	Active implant devices/ancillaries	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	10/12/2022	10/12/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	10/12/2022	10/12/2022
					CT2750	Implantable cardiac electrical stimulators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	10/12/2022	10/12/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	10/12/2022	10/12/2022
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	10/12/2022	10/12/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	10/12/2022	10/12/2022
					CT2985	Active	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	10/12/2022	10/12/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	10/12/2022	10/12/2022
					CT3001	Implantable metal	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	10/12/2022	10/12/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	10/12/2022	10/12/2022
					CT3002	Implantable synthetic polymer	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	10/12/2022	10/12/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	10/12/2022	10/12/2022
					CT3006	Implantable inorganic material	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	10/12/2022	10/12/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	10/12/2022	10/12/2022
					CT314	Electrically powered	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	10/12/2022	10/12/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	10/12/2022	10/12/2022
					CT321	Long-term surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	10/12/2022	10/12/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	10/12/2022	10/12/2022
					CT449	Generators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	10/12/2022	10/12/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	10/12/2022	10/12/2022
					CT981	Single use	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	10/12/2022	10/12/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	10/12/2022	10/12/2022
					CT983	Surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	10/12/2022	10/12/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	10/12/2022	10/12/2022
12093	2023		47265	Dual-chamber implantable pacemaker, rate-responsive	CT141	Electrodes/Leads and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	18/11/2022	18/11/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	18/11/2022	18/11/2022
					CT145	Cardiovascular devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	18/11/2022	18/11/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	18/11/2022	18/11/2022
					CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	18/11/2022	18/11/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	18/11/2022	18/11/2022
					CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	18/11/2022	18/11/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	18/11/2022	18/11/2022
					CT233	Surgical	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	18/11/2022	18/11/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	18/11/2022	18/11/2022
					CT2406	Implantable	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	18/11/2022	18/11/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	18/11/2022	18/11/2022
					CT2407	Active implantable	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	18/11/2022	18/11/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	18/11/2022	18/11/2022
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	18/11/2022	18/11/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	18/11/2022	18/11/2022
					CT2727	Active implant devices/ancillaries	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	18/11/2022	18/11/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	18/11/2022	18/11/2022
					CT2750	Implantable cardiac electrical stimulators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	18/11/2022	18/11/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	18/11/2022	18/11/2022
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	18/11/2022	18/11/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	18/11/2022	18/11/2022
					CT2985	Active	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	18/11/2022	18/11/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	18/11/2022	18/11/2022
					CT3001	Implantable metal	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	18/11/2022	18/11/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	18/11/2022	18/11/2022
					CT3002	Implantable synthetic polymer	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	18/11/2022	18/11/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	18/11/2022	18/11/2022
					CT3006	Implantable inorganic material	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	18/11/2022	18/11/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	18/11/2022	18/11/2022
					CT314	Electrically powered	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	18/11/2022	18/11/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	18/11/2022	18/11/2022

					CT321	Long-term surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	18/11/2022	18/11/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	18/11/2022	18/11/2022
					CT449	Generators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	18/11/2022	18/11/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	18/11/2022	18/11/2022
					CT981	Single use	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	18/11/2022	18/11/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	18/11/2022	18/11/2022
					CT983	Surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	18/11/2022	18/11/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	18/11/2022	18/11/2022
12094	2023		47265	Dual-chamber implantable pacemaker, rate-responsive	CT141	Electrodes/Leads and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	17/01/2023	17/01/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	17/01/2023	17/01/2023
					CT145	Cardiovascular devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	17/01/2023	17/01/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	17/01/2023	17/01/2023
					CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	17/01/2023	17/01/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	17/01/2023	17/01/2023
					CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	17/01/2023	17/01/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	17/01/2023	17/01/2023
					CT233	Surgical	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	17/01/2023	17/01/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	17/01/2023	17/01/2023
					CT2406	Implantable	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	17/01/2023	17/01/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	17/01/2023	17/01/2023
					CT2407	Active implantable	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	17/01/2023	17/01/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	17/01/2023	17/01/2023
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	17/01/2023	17/01/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	17/01/2023	17/01/2023
					CT2727	Active implant devices/ancillaries	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	17/01/2023	17/01/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	17/01/2023	17/01/2023
					CT2750	Implantable cardiac electrical stimulators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	17/01/2023	17/01/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	17/01/2023	17/01/2023
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	17/01/2023	17/01/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	17/01/2023	17/01/2023
					CT2985	Active	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	17/01/2023	17/01/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	17/01/2023	17/01/2023
					CT3001	Implantable metal	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	17/01/2023	17/01/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	17/01/2023	17/01/2023
					CT3002	Implantable synthetic polymer	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	17/01/2023	17/01/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	17/01/2023	17/01/2023
					CT3006	Implantable inorganic material	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	17/01/2023	17/01/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	17/01/2023	17/01/2023
					CT314	Electrically powered	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	17/01/2023	17/01/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	17/01/2023	17/01/2023
					CT321	Long-term surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	17/01/2023	17/01/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	17/01/2023	17/01/2023
					CT449	Generators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	17/01/2023	17/01/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	17/01/2023	17/01/2023
					CT981	Single use	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	17/01/2023	17/01/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	17/01/2023	17/01/2023
					CT983	Surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	17/01/2023	17/01/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	17/01/2023	17/01/2023
12095	2023	All other reportable incidents	47265	Dual-chamber implantable pacemaker, rate-responsive	CT141	Electrodes/Leads and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem		
							F1901	Additional Surgery	A24	No Apparent Device Problem		
					CT145	Cardiovascular devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem		
							F1901	Additional Surgery	A24	No Apparent Device Problem		
					CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem		
							F1901	Additional Surgery	A24	No Apparent Device Problem		
					CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem		
							F1901	Additional Surgery	A24	No Apparent Device Problem		
					CT233	Surgical	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem		
							F1901	Additional Surgery	A24	No Apparent Device Problem		
					CT2406	Implantable	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem		
							F1901	Additional Surgery	A24	No Apparent Device Problem		
					CT2407	Active implantable	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem		
							F1901	Additional Surgery	A24	No Apparent Device Problem		
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem		
							F1901	Additional Surgery	A24	No Apparent Device Problem		
					CT2727	Active implant devices/ancillaries	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem		
							F1901	Additional Surgery	A24	No Apparent Device Problem		
					CT2750	Implantable cardiac electrical stimulators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem		
							F1901	Additional Surgery	A24	No Apparent Device Problem		
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem		
							F1901	Additional Surgery	A24	No Apparent Device Problem		
					CT2985	Active	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem		
							F1901	Additional Surgery	A24	No Apparent Device Problem		
					CT3001	Implantable metal	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem		
							F1901	Additional Surgery	A24	No Apparent Device Problem		
					CT3002	Implantable synthetic polymer	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem		
							F1901	Additional Surgery	A24	No Apparent Device Problem		
					CT3006	Implantable inorganic material	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem		
							F1901	Additional Surgery	A24	No Apparent Device Problem		
					CT314	Electrically powered	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem		
							F1901	Additional Surgery	A24	No Apparent Device Problem		
					CT321	Long-term surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem		
							F1901	Additional Surgery	A24	No Apparent Device Problem		
					CT449	Generators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem		
							F1901	Additional Surgery	A24	No Apparent Device Problem		
					CT981	Single use	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem		
							F1901	Additional Surgery	A24	No Apparent Device Problem		
					CT983	Surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem		
							F1901	Additional Surgery	A24	No Apparent Device Problem		

12096	2023		47265	Dual-chamber implantable pacemaker, rate-responsive	CT141	Electrodes/Leads and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem		
							F1901	Additional Surgery	A24	No Apparent Device Problem		
					CT145	Cardiovascular devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem		
							F1901	Additional Surgery	A24	No Apparent Device Problem		
					CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem		
							F1901	Additional Surgery	A24	No Apparent Device Problem		
					CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem		
							F1901	Additional Surgery	A24	No Apparent Device Problem		
					CT233	Surgical	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem		
							F1901	Additional Surgery	A24	No Apparent Device Problem		
					CT2406	Implantable	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem		
							F1901	Additional Surgery	A24	No Apparent Device Problem		
					CT2407	Active implantable	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem		
							F1901	Additional Surgery	A24	No Apparent Device Problem		
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem		
							F1901	Additional Surgery	A24	No Apparent Device Problem		
					CT2727	Active implant devices/ancillaries	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem		
							F1901	Additional Surgery	A24	No Apparent Device Problem		
					CT2750	Implantable cardiac electrical stimulators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem		
							F1901	Additional Surgery	A24	No Apparent Device Problem		
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem		
							F1901	Additional Surgery	A24	No Apparent Device Problem		
					CT2985	Active	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem		
							F1901	Additional Surgery	A24	No Apparent Device Problem		
					CT3001	Implantable metal	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem		
							F1901	Additional Surgery	A24	No Apparent Device Problem		
					CT3002	Implantable synthetic polymer	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem		
							F1901	Additional Surgery	A24	No Apparent Device Problem		
					CT3006	Implantable inorganic material	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem		
							F1901	Additional Surgery	A24	No Apparent Device Problem		
					CT314	Electrically powered	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem		
							F1901	Additional Surgery	A24	No Apparent Device Problem		
					CT321	Long-term surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem		
							F1901	Additional Surgery	A24	No Apparent Device Problem		
					CT449	Generators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem		
							F1901	Additional Surgery	A24	No Apparent Device Problem		
					CT981	Single use	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem		
							F1901	Additional Surgery	A24	No Apparent Device Problem		
					CT983	Surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem		
							F1901	Additional Surgery	A24	No Apparent Device Problem		
12097	2023	All other reportable incidents	35223	Endocardial/interventricular septal pacing lead	CT1684	Cardiac pacemakers and associated devices	E0601	Arrhythmia	A070101	Failure to Capture		
							F26	No Health Consequences or Impact	A070101	Failure to Capture		
12098	2023		47265	Dual-chamber implantable pacemaker, rate-responsive	CT141	Electrodes/Leads and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	28/11/2022	28/11/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	28/11/2022	28/11/2022
					CT145	Cardiovascular devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	28/11/2022	28/11/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	28/11/2022	28/11/2022
					CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	28/11/2022	28/11/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	28/11/2022	28/11/2022
					CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	28/11/2022	28/11/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	28/11/2022	28/11/2022
					CT233	Surgical	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	28/11/2022	28/11/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	28/11/2022	28/11/2022
					CT2406	Implantable	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	28/11/2022	28/11/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	28/11/2022	28/11/2022
					CT2407	Active implantable	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	28/11/2022	28/11/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	28/11/2022	28/11/2022
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	28/11/2022	28/11/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	28/11/2022	28/11/2022
					CT2727	Active implant devices/ancillaries	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	28/11/2022	28/11/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	28/11/2022	28/11/2022
					CT2750	Implantable cardiac electrical stimulators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	28/11/2022	28/11/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	28/11/2022	28/11/2022
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	28/11/2022	28/11/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	28/11/2022	28/11/2022
					CT2985	Active	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	28/11/2022	28/11/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	28/11/2022	28/11/2022
					CT3001	Implantable metal	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	28/11/2022	28/11/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	28/11/2022	28/11/2022
					CT3002	Implantable synthetic polymer	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	28/11/2022	28/11/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	28/11/2022	28/11/2022
					CT3006	Implantable inorganic material	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	28/11/2022	28/11/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	28/11/2022	28/11/2022
					CT314	Electrically powered	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	28/11/2022	28/11/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	28/11/2022	28/11/2022
					CT321	Long-term surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	28/11/2022	28/11/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	28/11/2022	28/11/2022
					CT449	Generators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	28/11/2022	28/11/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	28/11/2022	28/11/2022
					CT981	Single use	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	28/11/2022	28/11/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	28/11/2022	28/11/2022
					CT983	Surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	28/11/2022	28/11/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	28/11/2022	28/11/2022
12099	2023		47265	Dual-chamber implantable pacemaker, rate-responsive	CT141	Electrodes/Leads and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	11/01/2023	11/01/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	11/01/2023	11/01/2023
					CT145	Cardiovascular devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	11/01/2023	11/01/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	11/01/2023	11/01/2023
					CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	11/01/2023	11/01/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	11/01/2023	11/01/2023

					CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	11/01/2023	11/01/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	11/01/2023	11/01/2023
					CT233	Surgical	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	11/01/2023	11/01/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	11/01/2023	11/01/2023
					CT2406	Implantable	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	11/01/2023	11/01/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	11/01/2023	11/01/2023
					CT2407	Active implantable	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	11/01/2023	11/01/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	11/01/2023	11/01/2023
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	11/01/2023	11/01/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	11/01/2023	11/01/2023
					CT2727	Active implant devices/ancillaries	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	11/01/2023	11/01/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	11/01/2023	11/01/2023
					CT2750	Implantable cardiac electrical stimulators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	11/01/2023	11/01/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	11/01/2023	11/01/2023
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	11/01/2023	11/01/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	11/01/2023	11/01/2023
					CT2985	Active	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	11/01/2023	11/01/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	11/01/2023	11/01/2023
					CT3001	Implantable metal	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	11/01/2023	11/01/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	11/01/2023	11/01/2023
					CT3002	Implantable synthetic polymer	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	11/01/2023	11/01/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	11/01/2023	11/01/2023
					CT3006	Implantable inorganic material	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	11/01/2023	11/01/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	11/01/2023	11/01/2023
					CT314	Electrically powered	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	11/01/2023	11/01/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	11/01/2023	11/01/2023
					CT321	Long-term surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	11/01/2023	11/01/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	11/01/2023	11/01/2023
					CT449	Generators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	11/01/2023	11/01/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	11/01/2023	11/01/2023
					CT981	Single use	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	11/01/2023	11/01/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	11/01/2023	11/01/2023
					CT983	Surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	11/01/2023	11/01/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	11/01/2023	11/01/2023
12100	2023		47265	Dual-chamber implantable pacemaker, rate-responsive	CT141	Electrodes/Leads and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	22/08/2022	22/08/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	22/08/2022	22/08/2022
					CT145	Cardiovascular devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	22/08/2022	22/08/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	22/08/2022	22/08/2022
					CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	22/08/2022	22/08/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	22/08/2022	22/08/2022
					CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	22/08/2022	22/08/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	22/08/2022	22/08/2022
					CT233	Surgical	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	22/08/2022	22/08/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	22/08/2022	22/08/2022
					CT2406	Implantable	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	22/08/2022	22/08/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	22/08/2022	22/08/2022
					CT2407	Active implantable	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	22/08/2022	22/08/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	22/08/2022	22/08/2022
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	22/08/2022	22/08/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	22/08/2022	22/08/2022
					CT2727	Active implant devices/ancillaries	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	22/08/2022	22/08/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	22/08/2022	22/08/2022
					CT2750	Implantable cardiac electrical stimulators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	22/08/2022	22/08/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	22/08/2022	22/08/2022
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	22/08/2022	22/08/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	22/08/2022	22/08/2022
					CT2985	Active	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	22/08/2022	22/08/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	22/08/2022	22/08/2022
					CT3001	Implantable metal	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	22/08/2022	22/08/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	22/08/2022	22/08/2022
					CT3002	Implantable synthetic polymer	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	22/08/2022	22/08/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	22/08/2022	22/08/2022
					CT3006	Implantable inorganic material	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	22/08/2022	22/08/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	22/08/2022	22/08/2022
					CT314	Electrically powered	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	22/08/2022	22/08/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	22/08/2022	22/08/2022
					CT321	Long-term surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	22/08/2022	22/08/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	22/08/2022	22/08/2022
					CT449	Generators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	22/08/2022	22/08/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	22/08/2022	22/08/2022
					CT981	Single use	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	22/08/2022	22/08/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	22/08/2022	22/08/2022
					CT983	Surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	22/08/2022	22/08/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	22/08/2022	22/08/2022
12101	2023		47265	Dual-chamber implantable pacemaker, rate-responsive	CT141	Electrodes/Leads and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	16/11/2022	16/11/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	16/11/2022	16/11/2022
					CT145	Cardiovascular devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	16/11/2022	16/11/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	16/11/2022	16/11/2022
					CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	16/11/2022	16/11/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	16/11/2022	16/11/2022
					CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	16/11/2022	16/11/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	16/11/2022	16/11/2022
					CT233	Surgical	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	16/11/2022	16/11/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	16/11/2022	16/11/2022
					CT2406	Implantable	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	16/11/2022	16/11/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	16/11/2022	16/11/2022
					CT2407	Active implantable	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	16/11/2022	16/11/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	16/11/2022	16/11/2022

					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	16/11/2022	16/11/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	16/11/2022	16/11/2022
					CT2727	Active implant devices/ancillaries	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	16/11/2022	16/11/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	16/11/2022	16/11/2022
					CT2750	Implantable cardiac electrical stimulators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	16/11/2022	16/11/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	16/11/2022	16/11/2022
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	16/11/2022	16/11/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	16/11/2022	16/11/2022
					CT2985	Active	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	16/11/2022	16/11/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	16/11/2022	16/11/2022
					CT3001	Implantable metal	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	16/11/2022	16/11/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	16/11/2022	16/11/2022
					CT3002	Implantable synthetic polymer	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	16/11/2022	16/11/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	16/11/2022	16/11/2022
					CT3006	Implantable inorganic material	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	16/11/2022	16/11/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	16/11/2022	16/11/2022
					CT314	Electrically powered	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	16/11/2022	16/11/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	16/11/2022	16/11/2022
					CT321	Long-term surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	16/11/2022	16/11/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	16/11/2022	16/11/2022
					CT449	Generators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	16/11/2022	16/11/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	16/11/2022	16/11/2022
					CT981	Single use	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	16/11/2022	16/11/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	16/11/2022	16/11/2022
					CT983	Surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	16/11/2022	16/11/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	16/11/2022	16/11/2022
12102	2023		47265	Dual-chamber implantable pacemaker, rate-responsive	CT141	Electrodes/Leads and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	31/10/2022	31/10/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	31/10/2022	31/10/2022
					CT145	Cardiovascular devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	31/10/2022	31/10/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	31/10/2022	31/10/2022
					CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	31/10/2022	31/10/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	31/10/2022	31/10/2022
					CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	31/10/2022	31/10/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	31/10/2022	31/10/2022
					CT233	Surgical	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	31/10/2022	31/10/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	31/10/2022	31/10/2022
					CT2406	Implantable	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	31/10/2022	31/10/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	31/10/2022	31/10/2022
					CT2407	Active implantable	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	31/10/2022	31/10/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	31/10/2022	31/10/2022
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	31/10/2022	31/10/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	31/10/2022	31/10/2022
					CT2727	Active implant devices/ancillaries	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	31/10/2022	31/10/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	31/10/2022	31/10/2022
					CT2750	Implantable cardiac electrical stimulators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	31/10/2022	31/10/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	31/10/2022	31/10/2022
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	31/10/2022	31/10/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	31/10/2022	31/10/2022
					CT2985	Active	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	31/10/2022	31/10/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	31/10/2022	31/10/2022
					CT3001	Implantable metal	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	31/10/2022	31/10/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	31/10/2022	31/10/2022
					CT3002	Implantable synthetic polymer	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	31/10/2022	31/10/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	31/10/2022	31/10/2022
					CT3006	Implantable inorganic material	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	31/10/2022	31/10/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	31/10/2022	31/10/2022
					CT314	Electrically powered	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	31/10/2022	31/10/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	31/10/2022	31/10/2022
					CT321	Long-term surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	31/10/2022	31/10/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	31/10/2022	31/10/2022
					CT449	Generators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	31/10/2022	31/10/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	31/10/2022	31/10/2022
					CT981	Single use	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	31/10/2022	31/10/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	31/10/2022	31/10/2022
					CT983	Surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	31/10/2022	31/10/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	31/10/2022	31/10/2022
12103	2023	All other reportable incidents	47265	Dual-chamber implantable pacemaker, rate-responsive	CT141	Electrodes/Leads and associated devices	E0112	Dizziness	A0711	Interrogation Problem	24/01/2023	24/01/2023
									A071102	Failure to Interrogate	24/01/2023	24/01/2023
									A0712	Pacing Problem	24/01/2023	24/01/2023
									A071204	No Pacing	24/01/2023	24/01/2023
							E0601	Arrhythmia	A0711	Interrogation Problem	24/01/2023	24/01/2023
									A071102	Failure to Interrogate	24/01/2023	24/01/2023
									A0712	Pacing Problem	24/01/2023	24/01/2023
									A071204	No Pacing	24/01/2023	24/01/2023
							E0717	Dyspnea	A0711	Interrogation Problem	24/01/2023	24/01/2023
									A071102	Failure to Interrogate	24/01/2023	24/01/2023
									A0712	Pacing Problem	24/01/2023	24/01/2023
									A071204	No Pacing	24/01/2023	24/01/2023
							F1901	Additional Surgery	A0711	Interrogation Problem	24/01/2023	24/01/2023
									A071102	Failure to Interrogate	24/01/2023	24/01/2023
									A0712	Pacing Problem	24/01/2023	24/01/2023
									A071204	No Pacing	24/01/2023	24/01/2023
					CT145	Cardiovascular devices	E0112	Dizziness	A0711	Interrogation Problem	24/01/2023	24/01/2023
									A071102	Failure to Interrogate	24/01/2023	24/01/2023
									A0712	Pacing Problem	24/01/2023	24/01/2023
									A071204	No Pacing	24/01/2023	24/01/2023
							E0601	Arrhythmia	A0711	Interrogation Problem	24/01/2023	24/01/2023
									A071102	Failure to Interrogate	24/01/2023	24/01/2023

									A0712	Pacing Problem	24/01/2023	24/01/2023
									A071204	No Pacing	24/01/2023	24/01/2023
							E0717	Dyspnea	A0711	Interrogation Problem	24/01/2023	24/01/2023
									A071102	Failure to Interrogate	24/01/2023	24/01/2023
									A0712	Pacing Problem	24/01/2023	24/01/2023
									A071204	No Pacing	24/01/2023	24/01/2023
							F1901	Additional Surgery	A0711	Interrogation Problem	24/01/2023	24/01/2023
									A071102	Failure to Interrogate	24/01/2023	24/01/2023
									A0712	Pacing Problem	24/01/2023	24/01/2023
									A071204	No Pacing	24/01/2023	24/01/2023
					CT1684	Cardiac pacemakers and associated devices	E0112	Dizziness	A0711	Interrogation Problem	24/01/2023	24/01/2023
									A071102	Failure to Interrogate	24/01/2023	24/01/2023
									A0712	Pacing Problem	24/01/2023	24/01/2023
									A071204	No Pacing	24/01/2023	24/01/2023
							E0601	Arrhythmia	A0711	Interrogation Problem	24/01/2023	24/01/2023
									A071102	Failure to Interrogate	24/01/2023	24/01/2023
									A0712	Pacing Problem	24/01/2023	24/01/2023
									A071204	No Pacing	24/01/2023	24/01/2023
							E0717	Dyspnea	A0711	Interrogation Problem	24/01/2023	24/01/2023
									A071102	Failure to Interrogate	24/01/2023	24/01/2023
									A0712	Pacing Problem	24/01/2023	24/01/2023
									A071204	No Pacing	24/01/2023	24/01/2023
							F1901	Additional Surgery	A0711	Interrogation Problem	24/01/2023	24/01/2023
									A071102	Failure to Interrogate	24/01/2023	24/01/2023
									A0712	Pacing Problem	24/01/2023	24/01/2023
									A071204	No Pacing	24/01/2023	24/01/2023
					CT1685	Cardiac pacemakers	E0112	Dizziness	A0711	Interrogation Problem	24/01/2023	24/01/2023
									A071102	Failure to Interrogate	24/01/2023	24/01/2023
									A0712	Pacing Problem	24/01/2023	24/01/2023
									A071204	No Pacing	24/01/2023	24/01/2023
							E0601	Arrhythmia	A0711	Interrogation Problem	24/01/2023	24/01/2023
									A071102	Failure to Interrogate	24/01/2023	24/01/2023
									A0712	Pacing Problem	24/01/2023	24/01/2023
									A071204	No Pacing	24/01/2023	24/01/2023
							E0717	Dyspnea	A0711	Interrogation Problem	24/01/2023	24/01/2023
									A071102	Failure to Interrogate	24/01/2023	24/01/2023
									A0712	Pacing Problem	24/01/2023	24/01/2023
									A071204	No Pacing	24/01/2023	24/01/2023
							F1901	Additional Surgery	A0711	Interrogation Problem	24/01/2023	24/01/2023
									A071102	Failure to Interrogate	24/01/2023	24/01/2023
									A0712	Pacing Problem	24/01/2023	24/01/2023
									A071204	No Pacing	24/01/2023	24/01/2023
					CT233	Surgical	E0112	Dizziness	A0711	Interrogation Problem	24/01/2023	24/01/2023
									A071102	Failure to Interrogate	24/01/2023	24/01/2023
									A0712	Pacing Problem	24/01/2023	24/01/2023
									A071204	No Pacing	24/01/2023	24/01/2023
							E0601	Arrhythmia	A0711	Interrogation Problem	24/01/2023	24/01/2023
									A071102	Failure to Interrogate	24/01/2023	24/01/2023
									A0712	Pacing Problem	24/01/2023	24/01/2023
									A071204	No Pacing	24/01/2023	24/01/2023
							E0717	Dyspnea	A0711	Interrogation Problem	24/01/2023	24/01/2023
									A071102	Failure to Interrogate	24/01/2023	24/01/2023
									A0712	Pacing Problem	24/01/2023	24/01/2023
									A071204	No Pacing	24/01/2023	24/01/2023
							F1901	Additional Surgery	A0711	Interrogation Problem	24/01/2023	24/01/2023
									A071102	Failure to Interrogate	24/01/2023	24/01/2023
									A0712	Pacing Problem	24/01/2023	24/01/2023
									A071204	No Pacing	24/01/2023	24/01/2023
					CT2406	Implantable	E0112	Dizziness	A0711	Interrogation Problem	24/01/2023	24/01/2023
									A071102	Failure to Interrogate	24/01/2023	24/01/2023
									A0712	Pacing Problem	24/01/2023	24/01/2023
									A071204	No Pacing	24/01/2023	24/01/2023
							E0601	Arrhythmia	A0711	Interrogation Problem	24/01/2023	24/01/2023
									A071102	Failure to Interrogate	24/01/2023	24/01/2023
									A0712	Pacing Problem	24/01/2023	24/01/2023
									A071204	No Pacing	24/01/2023	24/01/2023
							E0717	Dyspnea	A0711	Interrogation Problem	24/01/2023	24/01/2023
									A071102	Failure to Interrogate	24/01/2023	24/01/2023
									A0712	Pacing Problem	24/01/2023	24/01/2023
									A071204	No Pacing	24/01/2023	24/01/2023
							F1901	Additional Surgery	A0711	Interrogation Problem	24/01/2023	24/01/2023
									A071102	Failure to Interrogate	24/01/2023	24/01/2023
									A0712	Pacing Problem	24/01/2023	24/01/2023
									A071204	No Pacing	24/01/2023	24/01/2023
					CT2407	Active implantable	E0112	Dizziness	A0711	Interrogation Problem	24/01/2023	24/01/2023
									A071102	Failure to Interrogate	24/01/2023	24/01/2023
									A0712	Pacing Problem	24/01/2023	24/01/2023
									A071204	No Pacing	24/01/2023	24/01/2023
							E0601	Arrhythmia	A0711	Interrogation Problem	24/01/2023	24/01/2023
									A071102	Failure to Interrogate	24/01/2023	24/01/2023
									A0712	Pacing Problem	24/01/2023	24/01/2023
									A071204	No Pacing	24/01/2023	24/01/2023
							E0717	Dyspnea	A0711	Interrogation Problem	24/01/2023	24/01/2023
									A071102	Failure to Interrogate	24/01/2023	24/01/2023
									A0712	Pacing Problem	24/01/2023	24/01/2023
									A071204	No Pacing	24/01/2023	24/01/2023
							F1901	Additional Surgery	A0711	Interrogation Problem	24/01/2023	24/01/2023
									A071102	Failure to Interrogate	24/01/2023	24/01/2023

								A0712	Pacing Problem	24/01/2023	24/01/2023
								A071204	No Pacing	24/01/2023	24/01/2023
					CT257	Implantable cardiac pacemakers	E0112	Dizziness	A0711	Interrogation Problem	24/01/2023	24/01/2023
									A071102	Failure to Interrogate	24/01/2023	24/01/2023
									A0712	Pacing Problem	24/01/2023	24/01/2023
									A071204	No Pacing	24/01/2023	24/01/2023
							E0601	Arrhythmia	A0711	Interrogation Problem	24/01/2023	24/01/2023
									A071102	Failure to Interrogate	24/01/2023	24/01/2023
									A0712	Pacing Problem	24/01/2023	24/01/2023
									A071204	No Pacing	24/01/2023	24/01/2023
							E0717	Dyspnea	A0711	Interrogation Problem	24/01/2023	24/01/2023
									A071102	Failure to Interrogate	24/01/2023	24/01/2023
									A0712	Pacing Problem	24/01/2023	24/01/2023
									A071204	No Pacing	24/01/2023	24/01/2023
							F1901	Additional Surgery	A0711	Interrogation Problem	24/01/2023	24/01/2023
									A071102	Failure to Interrogate	24/01/2023	24/01/2023
									A0712	Pacing Problem	24/01/2023	24/01/2023
									A071204	No Pacing	24/01/2023	24/01/2023
					CT2727	Active implant devices/ancillaries	E0112	Dizziness	A0711	Interrogation Problem	24/01/2023	24/01/2023
									A071102	Failure to Interrogate	24/01/2023	24/01/2023
									A0712	Pacing Problem	24/01/2023	24/01/2023
									A071204	No Pacing	24/01/2023	24/01/2023
							E0601	Arrhythmia	A0711	Interrogation Problem	24/01/2023	24/01/2023
									A071102	Failure to Interrogate	24/01/2023	24/01/2023
									A0712	Pacing Problem	24/01/2023	24/01/2023
									A071204	No Pacing	24/01/2023	24/01/2023
							E0717	Dyspnea	A0711	Interrogation Problem	24/01/2023	24/01/2023
									A071102	Failure to Interrogate	24/01/2023	24/01/2023
									A0712	Pacing Problem	24/01/2023	24/01/2023
									A071204	No Pacing	24/01/2023	24/01/2023
							F1901	Additional Surgery	A0711	Interrogation Problem	24/01/2023	24/01/2023
									A071102	Failure to Interrogate	24/01/2023	24/01/2023
									A0712	Pacing Problem	24/01/2023	24/01/2023
									A071204	No Pacing	24/01/2023	24/01/2023
					CT2750	Implantable cardiac electrical stimulators and associated devices	E0112	Dizziness	A0711	Interrogation Problem	24/01/2023	24/01/2023
									A071102	Failure to Interrogate	24/01/2023	24/01/2023
									A0712	Pacing Problem	24/01/2023	24/01/2023
									A071204	No Pacing	24/01/2023	24/01/2023
							E0601	Arrhythmia	A0711	Interrogation Problem	24/01/2023	24/01/2023
									A071102	Failure to Interrogate	24/01/2023	24/01/2023
									A0712	Pacing Problem	24/01/2023	24/01/2023
									A071204	No Pacing	24/01/2023	24/01/2023
							E0717	Dyspnea	A0711	Interrogation Problem	24/01/2023	24/01/2023
									A071102	Failure to Interrogate	24/01/2023	24/01/2023
									A0712	Pacing Problem	24/01/2023	24/01/2023
									A071204	No Pacing	24/01/2023	24/01/2023
							F1901	Additional Surgery	A0711	Interrogation Problem	24/01/2023	24/01/2023
									A071102	Failure to Interrogate	24/01/2023	24/01/2023
									A0712	Pacing Problem	24/01/2023	24/01/2023
									A071204	No Pacing	24/01/2023	24/01/2023
					CT2866	Implantable cardiac electrical stimulators	E0112	Dizziness	A0711	Interrogation Problem	24/01/2023	24/01/2023
									A071102	Failure to Interrogate	24/01/2023	24/01/2023
									A0712	Pacing Problem	24/01/2023	24/01/2023
									A071204	No Pacing	24/01/2023	24/01/2023
							E0601	Arrhythmia	A0711	Interrogation Problem	24/01/2023	24/01/2023
									A071102	Failure to Interrogate	24/01/2023	24/01/2023
									A0712	Pacing Problem	24/01/2023	24/01/2023
									A071204	No Pacing	24/01/2023	24/01/2023
							E0717	Dyspnea	A0711	Interrogation Problem	24/01/2023	24/01/2023
									A071102	Failure to Interrogate	24/01/2023	24/01/2023
									A0712	Pacing Problem	24/01/2023	24/01/2023
									A071204	No Pacing	24/01/2023	24/01/2023
							F1901	Additional Surgery	A0711	Interrogation Problem	24/01/2023	24/01/2023
									A071102	Failure to Interrogate	24/01/2023	24/01/2023
									A0712	Pacing Problem	24/01/2023	24/01/2023
									A071204	No Pacing	24/01/2023	24/01/2023
					CT2985	Active	E0112	Dizziness	A0711	Interrogation Problem	24/01/2023	24/01/2023
									A071102	Failure to Interrogate	24/01/2023	24/01/2023
									A0712	Pacing Problem	24/01/2023	24/01/2023
									A071204	No Pacing	24/01/2023	24/01/2023
							E0601	Arrhythmia	A0711	Interrogation Problem	24/01/2023	24/01/2023
									A071102	Failure to Interrogate	24/01/2023	24/01/2023
									A0712	Pacing Problem	24/01/2023	24/01/2023
									A071204	No Pacing	24/01/2023	24/01/2023
							E0717	Dyspnea	A0711	Interrogation Problem	24/01/2023	24/01/2023
									A071102	Failure to Interrogate	24/01/2023	24/01/2023
									A0712	Pacing Problem	24/01/2023	24/01/2023
									A071204	No Pacing	24/01/2023	24/01/2023
							F1901	Additional Surgery	A0711	Interrogation Problem	24/01/2023	24/01/2023
									A071102	Failure to Interrogate	24/01/2023	24/01/2023
									A0712	Pacing Problem	24/01/2023	24/01/2023
									A071204	No Pacing	24/01/2023	24/01/2023
					CT3001	Implantable metal	E0112	Dizziness	A0711	Interrogation Problem	24/01/2023	24/01/2023
									A071102	Failure to Interrogate	24/01/2023	24/01/2023
									A0712	Pacing Problem	24/01/2023	24/01/2023
									A071204	No Pacing	24/01/2023	24/01/2023
							E0601	Arrhythmia	A0711	Interrogation Problem	24/01/2023	24/01/2023
									A071102	Failure to Interrogate	24/01/2023	24/01/2023

								A0712	Pacing Problem	24/01/2023	24/01/2023
								A071204	No Pacing	24/01/2023	24/01/2023
						E0717	Dyspnea	A0711	Interrogation Problem	24/01/2023	24/01/2023
								A071102	Failure to Interrogate	24/01/2023	24/01/2023
								A0712	Pacing Problem	24/01/2023	24/01/2023
								A071204	No Pacing	24/01/2023	24/01/2023
						F1901	Additional Surgery	A0711	Interrogation Problem	24/01/2023	24/01/2023
								A071102	Failure to Interrogate	24/01/2023	24/01/2023
								A0712	Pacing Problem	24/01/2023	24/01/2023
								A071204	No Pacing	24/01/2023	24/01/2023
					CT3002	Implantable synthetic polymer	E0112	Dizziness	A0711	Interrogation Problem	24/01/2023	24/01/2023
									A071102	Failure to Interrogate	24/01/2023	24/01/2023
									A0712	Pacing Problem	24/01/2023	24/01/2023
									A071204	No Pacing	24/01/2023	24/01/2023
							E0601	Arrhythmia	A0711	Interrogation Problem	24/01/2023	24/01/2023
									A071102	Failure to Interrogate	24/01/2023	24/01/2023
									A0712	Pacing Problem	24/01/2023	24/01/2023
									A071204	No Pacing	24/01/2023	24/01/2023
							E0717	Dyspnea	A0711	Interrogation Problem	24/01/2023	24/01/2023
									A071102	Failure to Interrogate	24/01/2023	24/01/2023
									A0712	Pacing Problem	24/01/2023	24/01/2023
									A071204	No Pacing	24/01/2023	24/01/2023
							F1901	Additional Surgery	A0711	Interrogation Problem	24/01/2023	24/01/2023
									A071102	Failure to Interrogate	24/01/2023	24/01/2023
									A0712	Pacing Problem	24/01/2023	24/01/2023
									A071204	No Pacing	24/01/2023	24/01/2023
					CT3006	Implantable inorganic material	E0112	Dizziness	A0711	Interrogation Problem	24/01/2023	24/01/2023
									A071102	Failure to Interrogate	24/01/2023	24/01/2023
									A0712	Pacing Problem	24/01/2023	24/01/2023
									A071204	No Pacing	24/01/2023	24/01/2023
							E0601	Arrhythmia	A0711	Interrogation Problem	24/01/2023	24/01/2023
									A071102	Failure to Interrogate	24/01/2023	24/01/2023
									A0712	Pacing Problem	24/01/2023	24/01/2023
									A071204	No Pacing	24/01/2023	24/01/2023
							E0717	Dyspnea	A0711	Interrogation Problem	24/01/2023	24/01/2023
									A071102	Failure to Interrogate	24/01/2023	24/01/2023
									A0712	Pacing Problem	24/01/2023	24/01/2023
									A071204	No Pacing	24/01/2023	24/01/2023
							F1901	Additional Surgery	A0711	Interrogation Problem	24/01/2023	24/01/2023
									A071102	Failure to Interrogate	24/01/2023	24/01/2023
									A0712	Pacing Problem	24/01/2023	24/01/2023
									A071204	No Pacing	24/01/2023	24/01/2023
					CT314	Electrically powered	E0112	Dizziness	A0711	Interrogation Problem	24/01/2023	24/01/2023
									A071102	Failure to Interrogate	24/01/2023	24/01/2023
									A0712	Pacing Problem	24/01/2023	24/01/2023
									A071204	No Pacing	24/01/2023	24/01/2023
							E0601	Arrhythmia	A0711	Interrogation Problem	24/01/2023	24/01/2023
									A071102	Failure to Interrogate	24/01/2023	24/01/2023
									A0712	Pacing Problem	24/01/2023	24/01/2023
									A071204	No Pacing	24/01/2023	24/01/2023
							E0717	Dyspnea	A0711	Interrogation Problem	24/01/2023	24/01/2023
									A071102	Failure to Interrogate	24/01/2023	24/01/2023
									A0712	Pacing Problem	24/01/2023	24/01/2023
									A071204	No Pacing	24/01/2023	24/01/2023
							F1901	Additional Surgery	A0711	Interrogation Problem	24/01/2023	24/01/2023
									A071102	Failure to Interrogate	24/01/2023	24/01/2023
									A0712	Pacing Problem	24/01/2023	24/01/2023
									A071204	No Pacing	24/01/2023	24/01/2023
					CT321	Long-term surgical invasive	E0112	Dizziness	A0711	Interrogation Problem	24/01/2023	24/01/2023
									A071102	Failure to Interrogate	24/01/2023	24/01/2023
									A0712	Pacing Problem	24/01/2023	24/01/2023
									A071204	No Pacing	24/01/2023	24/01/2023
							E0601	Arrhythmia	A0711	Interrogation Problem	24/01/2023	24/01/2023
									A071102	Failure to Interrogate	24/01/2023	24/01/2023
									A0712	Pacing Problem	24/01/2023	24/01/2023
									A071204	No Pacing	24/01/2023	24/01/2023
							E0717	Dyspnea	A0711	Interrogation Problem	24/01/2023	24/01/2023
									A071102	Failure to Interrogate	24/01/2023	24/01/2023
									A0712	Pacing Problem	24/01/2023	24/01/2023
									A071204	No Pacing	24/01/2023	24/01/2023
							F1901	Additional Surgery	A0711	Interrogation Problem	24/01/2023	24/01/2023
									A071102	Failure to Interrogate	24/01/2023	24/01/2023
									A0712	Pacing Problem	24/01/2023	24/01/2023
									A071204	No Pacing	24/01/2023	24/01/2023
					CT449	Generators and associated devices	E0112	Dizziness	A0711	Interrogation Problem	24/01/2023	24/01/2023
									A071102	Failure to Interrogate	24/01/2023	24/01/2023
									A0712	Pacing Problem	24/01/2023	24/01/2023
									A071204	No Pacing	24/01/2023	24/01/2023
							E0601	Arrhythmia	A0711	Interrogation Problem	24/01/2023	24/01/2023
									A071102	Failure to Interrogate	24/01/2023	24/01/2023
									A0712	Pacing Problem	24/01/2023	24/01/2023
									A071204	No Pacing	24/01/2023	24/01/2023
							E0717	Dyspnea	A0711	Interrogation Problem	24/01/2023	24/01/2023
									A071102	Failure to Interrogate	24/01/2023	24/01/2023
									A0712	Pacing Problem	24/01/2023	24/01/2023
									A071204	No Pacing	24/01/2023	24/01/2023
							F1901	Additional Surgery	A0711	Interrogation Problem	24/01/2023	24/01/2023
									A071102	Failure to Interrogate	24/01/2023	24/01/2023

								A0712	Pacing Problem	24/01/2023	24/01/2023
								A071204	No Pacing	24/01/2023	24/01/2023
					CT981	Single use	E0112	Dizziness	A0711	Interrogation Problem	24/01/2023	24/01/2023
									A071102	Failure to Interrogate	24/01/2023	24/01/2023
									A0712	Pacing Problem	24/01/2023	24/01/2023
									A071204	No Pacing	24/01/2023	24/01/2023
							E0601	Arrhythmia	A0711	Interrogation Problem	24/01/2023	24/01/2023
									A071102	Failure to Interrogate	24/01/2023	24/01/2023
									A0712	Pacing Problem	24/01/2023	24/01/2023
									A071204	No Pacing	24/01/2023	24/01/2023
							E0717	Dyspnea	A0711	Interrogation Problem	24/01/2023	24/01/2023
									A071102	Failure to Interrogate	24/01/2023	24/01/2023
									A0712	Pacing Problem	24/01/2023	24/01/2023
									A071204	No Pacing	24/01/2023	24/01/2023
							F1901	Additional Surgery	A0711	Interrogation Problem	24/01/2023	24/01/2023
									A071102	Failure to Interrogate	24/01/2023	24/01/2023
									A0712	Pacing Problem	24/01/2023	24/01/2023
									A071204	No Pacing	24/01/2023	24/01/2023
					CT983	Surgical invasive	E0112	Dizziness	A0711	Interrogation Problem	24/01/2023	24/01/2023
									A071102	Failure to Interrogate	24/01/2023	24/01/2023
									A0712	Pacing Problem	24/01/2023	24/01/2023
									A071204	No Pacing	24/01/2023	24/01/2023
							E0601	Arrhythmia	A0711	Interrogation Problem	24/01/2023	24/01/2023
									A071102	Failure to Interrogate	24/01/2023	24/01/2023
									A0712	Pacing Problem	24/01/2023	24/01/2023
									A071204	No Pacing	24/01/2023	24/01/2023
							E0717	Dyspnea	A0711	Interrogation Problem	24/01/2023	24/01/2023
									A071102	Failure to Interrogate	24/01/2023	24/01/2023
									A0712	Pacing Problem	24/01/2023	24/01/2023
									A071204	No Pacing	24/01/2023	24/01/2023
							F1901	Additional Surgery	A0711	Interrogation Problem	24/01/2023	24/01/2023
									A071102	Failure to Interrogate	24/01/2023	24/01/2023
									A0712	Pacing Problem	24/01/2023	24/01/2023
									A071204	No Pacing	24/01/2023	24/01/2023
12104	2023	All other reportable incidents	47265	Dual-chamber implantable pacemaker, rate-responsive	CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A0908	Incorrect, Inadequate or Imprecise Result or Readings		
							F26	No Health Consequences or Impact	A0908	Incorrect, Inadequate or Imprecise Result or Readings		
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A0908	Incorrect, Inadequate or Imprecise Result or Readings		
							F26	No Health Consequences or Impact	A0908	Incorrect, Inadequate or Imprecise Result or Readings		
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A0908	Incorrect, Inadequate or Imprecise Result or Readings		
							F26	No Health Consequences or Impact	A0908	Incorrect, Inadequate or Imprecise Result or Readings		
12105	2023	All other reportable incidents	47263	Cardiac resynchronization therapy implantable pacemaker	CT1685	Cardiac pacemakers	E2103	Device Overstimulation of Tissue	A0908	Incorrect, Inadequate or Imprecise Result or Readings	16/01/2023	16/01/2023
							F08	Hospitalization or Prolonged Hospitalization	A0908	Incorrect, Inadequate or Imprecise Result or Readings	16/01/2023	16/01/2023
							F1905	Device Revision or Replacement	A0908	Incorrect, Inadequate or Imprecise Result or Readings	16/01/2023	16/01/2023
					CT257	Implantable cardiac pacemakers	E2103	Device Overstimulation of Tissue	A0908	Incorrect, Inadequate or Imprecise Result or Readings	16/01/2023	16/01/2023
							F08	Hospitalization or Prolonged Hospitalization	A0908	Incorrect, Inadequate or Imprecise Result or Readings	16/01/2023	16/01/2023
							F1905	Device Revision or Replacement	A0908	Incorrect, Inadequate or Imprecise Result or Readings	16/01/2023	16/01/2023
					CT2866	Implantable cardiac electrical stimulators	E2103	Device Overstimulation of Tissue	A0908	Incorrect, Inadequate or Imprecise Result or Readings	16/01/2023	16/01/2023
							F08	Hospitalization or Prolonged Hospitalization	A0908	Incorrect, Inadequate or Imprecise Result or Readings	16/01/2023	16/01/2023
							F1905	Device Revision or Replacement	A0908	Incorrect, Inadequate or Imprecise Result or Readings	16/01/2023	16/01/2023
12106	2023		47265	Dual-chamber implantable pacemaker, rate-responsive	CT141	Electrodes/Leads and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	06/12/2022	06/12/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	06/12/2022	06/12/2022
					CT145	Cardiovascular devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	06/12/2022	06/12/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	06/12/2022	06/12/2022
					CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	06/12/2022	06/12/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	06/12/2022	06/12/2022
					CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	06/12/2022	06/12/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	06/12/2022	06/12/2022
					CT233	Surgical	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	06/12/2022	06/12/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	06/12/2022	06/12/2022
					CT2406	Implantable	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	06/12/2022	06/12/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	06/12/2022	06/12/2022
					CT2407	Active implantable	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	06/12/2022	06/12/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	06/12/2022	06/12/2022
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	06/12/2022	06/12/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	06/12/2022	06/12/2022
					CT2727	Active implant devices/ancillaries	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	06/12/2022	06/12/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	06/12/2022	06/12/2022
					CT2750	Implantable cardiac electrical stimulators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	06/12/2022	06/12/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	06/12/2022	06/12/2022
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	06/12/2022	06/12/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	06/12/2022	06/12/2022
					CT2985	Active	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	06/12/2022	06/12/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	06/12/2022	06/12/2022
					CT3001	Implantable metal	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	06/12/2022	06/12/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	06/12/2022	06/12/2022
					CT3002	Implantable synthetic polymer	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	06/12/2022	06/12/2022

						F1901	Additional Surgery	A24	No Apparent Device Problem	06/12/2022	06/12/2022
				CT3006	Implantable inorganic material	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	06/12/2022	06/12/2022
						F1901	Additional Surgery	A24	No Apparent Device Problem	06/12/2022	06/12/2022
				CT314	Electrically powered	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	06/12/2022	06/12/2022
						F1901	Additional Surgery	A24	No Apparent Device Problem	06/12/2022	06/12/2022
				CT321	Long-term surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	06/12/2022	06/12/2022
						F1901	Additional Surgery	A24	No Apparent Device Problem	06/12/2022	06/12/2022
				CT449	Generators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	06/12/2022	06/12/2022
						F1901	Additional Surgery	A24	No Apparent Device Problem	06/12/2022	06/12/2022
				CT981	Single use	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	06/12/2022	06/12/2022
						F1901	Additional Surgery	A24	No Apparent Device Problem	06/12/2022	06/12/2022
				CT983	Surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	06/12/2022	06/12/2022
						F1901	Additional Surgery	A24	No Apparent Device Problem	06/12/2022	06/12/2022
12107	2023	All other reportable incidents	47265	Dual-chamber implantable pacemaker, rate-responsive	CT141	Electrodes/Leads and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem		
							F1901	Additional Surgery	A24	No Apparent Device Problem		
					CT145	Cardiovascular devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem		
							F1901	Additional Surgery	A24	No Apparent Device Problem		
					CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem		
							F1901	Additional Surgery	A24	No Apparent Device Problem		
					CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem		
							F1901	Additional Surgery	A24	No Apparent Device Problem		
					CT233	Surgical	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem		
							F1901	Additional Surgery	A24	No Apparent Device Problem		
					CT2406	Implantable	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem		
							F1901	Additional Surgery	A24	No Apparent Device Problem		
					CT2407	Active implantable	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem		
							F1901	Additional Surgery	A24	No Apparent Device Problem		
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem		
							F1901	Additional Surgery	A24	No Apparent Device Problem		
					CT2727	Active implant devices/ancillaries	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem		
							F1901	Additional Surgery	A24	No Apparent Device Problem		
					CT2750	Implantable cardiac electrical stimulators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem		
							F1901	Additional Surgery	A24	No Apparent Device Problem		
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem		
							F1901	Additional Surgery	A24	No Apparent Device Problem		
					CT2985	Active	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem		
							F1901	Additional Surgery	A24	No Apparent Device Problem		
					CT3001	Implantable metal	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem		
							F1901	Additional Surgery	A24	No Apparent Device Problem		
					CT3002	Implantable synthetic polymer	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem		
							F1901	Additional Surgery	A24	No Apparent Device Problem		
					CT3006	Implantable inorganic material	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem		
							F1901	Additional Surgery	A24	No Apparent Device Problem		
					CT314	Electrically powered	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem		
							F1901	Additional Surgery	A24	No Apparent Device Problem		
					CT321	Long-term surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem		
							F1901	Additional Surgery	A24	No Apparent Device Problem		
					CT449	Generators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem		
							F1901	Additional Surgery	A24	No Apparent Device Problem		
					CT981	Single use	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem		
							F1901	Additional Surgery	A24	No Apparent Device Problem		
					CT983	Surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem		
							F1901	Additional Surgery	A24	No Apparent Device Problem		
12108	2023		47265	Dual-chamber implantable pacemaker, rate-responsive	CT141	Electrodes/Leads and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A090805	Incorrect Measurement		
							F26	No Health Consequences or Impact	A090805	Incorrect Measurement		
					CT145	Cardiovascular devices	E2403	No Clinical Signs, Symptoms or Conditions	A090805	Incorrect Measurement		
							F26	No Health Consequences or Impact	A090805	Incorrect Measurement		
					CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A090805	Incorrect Measurement		
							F26	No Health Consequences or Impact	A090805	Incorrect Measurement		
					CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A090805	Incorrect Measurement		
							F26	No Health Consequences or Impact	A090805	Incorrect Measurement		
					CT233	Surgical	E2403	No Clinical Signs, Symptoms or Conditions	A090805	Incorrect Measurement		
							F26	No Health Consequences or Impact	A090805	Incorrect Measurement		
					CT2406	Implantable	E2403	No Clinical Signs, Symptoms or Conditions	A090805	Incorrect Measurement		
							F26	No Health Consequences or Impact	A090805	Incorrect Measurement		
					CT2407	Active implantable	E2403	No Clinical Signs, Symptoms or Conditions	A090805	Incorrect Measurement		
							F26	No Health Consequences or Impact	A090805	Incorrect Measurement		
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A090805	Incorrect Measurement		
							F26	No Health Consequences or Impact	A090805	Incorrect Measurement		
					CT2727	Active implant devices/ancillaries	E2403	No Clinical Signs, Symptoms or Conditions	A090805	Incorrect Measurement		
							F26	No Health Consequences or Impact	A090805	Incorrect Measurement		
					CT2750	Implantable cardiac electrical stimulators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A090805	Incorrect Measurement		
							F26	No Health Consequences or Impact	A090805	Incorrect Measurement		
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A090805	Incorrect Measurement		
							F26	No Health Consequences or Impact	A090805	Incorrect Measurement		
					CT2985	Active	E2403	No Clinical Signs, Symptoms or Conditions	A090805	Incorrect Measurement		
							F26	No Health Consequences or Impact	A090805	Incorrect Measurement		
					CT3001	Implantable metal	E2403	No Clinical Signs, Symptoms or Conditions	A090805	Incorrect Measurement		
							F26	No Health Consequences or Impact	A090805	Incorrect Measurement		
					CT3002	Implantable synthetic polymer	E2403	No Clinical Signs, Symptoms or Conditions	A090805	Incorrect Measurement		
							F26	No Health Consequences or Impact	A090805	Incorrect Measurement		
					CT3006	Implantable inorganic material	E2403	No Clinical Signs, Symptoms or Conditions	A090805	Incorrect Measurement		
							F26	No Health Consequences or Impact	A090805	Incorrect Measurement		
					CT314	Electrically powered	E2403	No Clinical Signs, Symptoms or Conditions	A090805	Incorrect Measurement		
							F26	No Health Consequences or Impact	A090805	Incorrect Measurement		
					CT321	Long-term surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A090805	Incorrect Measurement		
							F26	No Health Consequences or Impact	A090805	Incorrect Measurement		
					CT449	Generators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A090805	Incorrect Measurement		

							F26	No Health Consequences or Impact	A090805	Incorrect Measurement			
					CT981	Single use	E2403	No Clinical Signs, Symptoms or Conditions	A090805	Incorrect Measurement			
							F26	No Health Consequences or Impact	A090805	Incorrect Measurement			
					CT983	Surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A090805	Incorrect Measurement			
							F26	No Health Consequences or Impact	A090805	Incorrect Measurement			
12109	2023		47265	Dual-chamber implantable pacemaker, rate-responsive	CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	14/12/2022	14/12/2022	
							F1901	Additional Surgery	A24	No Apparent Device Problem	14/12/2022	14/12/2022	
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	14/12/2022	14/12/2022	
							F1901	Additional Surgery	A24	No Apparent Device Problem	14/12/2022	14/12/2022	
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	14/12/2022	14/12/2022	
							F1901	Additional Surgery	A24	No Apparent Device Problem	14/12/2022	14/12/2022	
12110	2023	All other reportable incidents	47265	Dual-chamber implantable pacemaker, rate-responsive	CT141	Electrodes/Leads and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem			
							F1901	Additional Surgery	A24	No Apparent Device Problem			
					CT145	Cardiovascular devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem			
							F1901	Additional Surgery	A24	No Apparent Device Problem			
					CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem			
							F1901	Additional Surgery	A24	No Apparent Device Problem			
					CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem			
							F1901	Additional Surgery	A24	No Apparent Device Problem			
					CT233	Surgical	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem			
							F1901	Additional Surgery	A24	No Apparent Device Problem			
					CT2406	Implantable	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem			
							F1901	Additional Surgery	A24	No Apparent Device Problem			
					CT2407	Active implantable	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem			
							F1901	Additional Surgery	A24	No Apparent Device Problem			
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem			
							F1901	Additional Surgery	A24	No Apparent Device Problem			
					CT2727	Active implant devices/ancillaries	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem			
							F1901	Additional Surgery	A24	No Apparent Device Problem			
					CT2750	Implantable cardiac electrical stimulators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem			
							F1901	Additional Surgery	A24	No Apparent Device Problem			
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem			
							F1901	Additional Surgery	A24	No Apparent Device Problem			
					CT2985	Active	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem			
							F1901	Additional Surgery	A24	No Apparent Device Problem			
					CT3001	Implantable metal	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem			
							F1901	Additional Surgery	A24	No Apparent Device Problem			
					CT3002	Implantable synthetic polymer	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem			
							F1901	Additional Surgery	A24	No Apparent Device Problem			
					CT3006	Implantable inorganic material	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem			
							F1901	Additional Surgery	A24	No Apparent Device Problem			
					CT314	Electrically powered	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem			
							F1901	Additional Surgery	A24	No Apparent Device Problem			
					CT321	Long-term surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem			
							F1901	Additional Surgery	A24	No Apparent Device Problem			
					CT449	Generators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem			
							F1901	Additional Surgery	A24	No Apparent Device Problem			
					CT981	Single use	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem			
							F1901	Additional Surgery	A24	No Apparent Device Problem			
					CT983	Surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem			
							F1901	Additional Surgery	A24	No Apparent Device Problem			
12111	2023	All other reportable incidents	47265	Dual-chamber implantable pacemaker, rate-responsive	CT141	Electrodes/Leads and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	28/11/2022	28/11/2022	
							F1901	Additional Surgery	A24	No Apparent Device Problem	28/11/2022	28/11/2022	
					CT145	Cardiovascular devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	28/11/2022	28/11/2022	
							F1901	Additional Surgery	A24	No Apparent Device Problem	28/11/2022	28/11/2022	
					CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	28/11/2022	28/11/2022	
							F1901	Additional Surgery	A24	No Apparent Device Problem	28/11/2022	28/11/2022	
					CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	28/11/2022	28/11/2022	
							F1901	Additional Surgery	A24	No Apparent Device Problem	28/11/2022	28/11/2022	
					CT233	Surgical	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	28/11/2022	28/11/2022	
							F1901	Additional Surgery	A24	No Apparent Device Problem	28/11/2022	28/11/2022	
					CT2406	Implantable	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	28/11/2022	28/11/2022	
							F1901	Additional Surgery	A24	No Apparent Device Problem	28/11/2022	28/11/2022	
					CT2407	Active implantable	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	28/11/2022	28/11/2022	
							F1901	Additional Surgery	A24	No Apparent Device Problem	28/11/2022	28/11/2022	
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	28/11/2022	28/11/2022	
							F1901	Additional Surgery	A24	No Apparent Device Problem	28/11/2022	28/11/2022	
					CT2727	Active implant devices/ancillaries	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	28/11/2022	28/11/2022	
							F1901	Additional Surgery	A24	No Apparent Device Problem	28/11/2022	28/11/2022	
					CT2750	Implantable cardiac electrical stimulators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	28/11/2022	28/11/2022	
							F1901	Additional Surgery	A24	No Apparent Device Problem	28/11/2022	28/11/2022	
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	28/11/2022	28/11/2022	
							F1901	Additional Surgery	A24	No Apparent Device Problem	28/11/2022	28/11/2022	
					CT2985	Active	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	28/11/2022	28/11/2022	
							F1901	Additional Surgery	A24	No Apparent Device Problem	28/11/2022	28/11/2022	
					CT3001	Implantable metal	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	28/11/2022	28/11/2022	
							F1901	Additional Surgery	A24	No Apparent Device Problem	28/11/2022	28/11/2022	
					CT3002	Implantable synthetic polymer	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	28/11/2022	28/11/2022	
							F1901	Additional Surgery	A24	No Apparent Device Problem	28/11/2022	28/11/2022	
					CT3006	Implantable inorganic material	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	28/11/2022	28/11/2022	
							F1901	Additional Surgery	A24	No Apparent Device Problem	28/11/2022	28/11/2022	
					CT314	Electrically powered	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	28/11/2022	28/11/2022	
							F1901	Additional Surgery	A24	No Apparent Device Problem	28/11/2022	28/11/2022	
					CT321	Long-term surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	28/11/2022	28/11/2022	
							F1901	Additional Surgery	A24	No Apparent Device Problem	28/11/2022	28/11/2022	
					CT449	Generators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	28/11/2022	28/11/2022	
							F1901	Additional Surgery	A24	No Apparent Device Problem	28/11/2022	28/11/2022	
					CT981	Single use	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	28/11/2022	28/11/2022	

						F1901	Additional Surgery	A24	No Apparent Device Problem	28/11/2022	28/11/2022
					CT983	Surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	28/11/2022	28/11/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	28/11/2022	28/11/2022
12112	2023		47265	Dual-chamber implantable pacemaker, rate-responsive	CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	13/12/2022	13/12/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	13/12/2022	13/12/2022
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	13/12/2022	13/12/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	13/12/2022	13/12/2022
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	13/12/2022	13/12/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	13/12/2022	13/12/2022
12113	2023	All other reportable incidents	36241	Pacemaker/defibrillator lead	CT1684	Cardiac pacemakers and associated devices	E2104	Electric Shock	A071302	Inappropriate/Inadequate Shock/Stimulation		
							F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A071302	Inappropriate/Inadequate Shock/Stimulation		
12114	2023		47265	Dual-chamber implantable pacemaker, rate-responsive	CT141	Electrodes/Leads and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	22/12/2022	22/12/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	22/12/2022	22/12/2022
					CT145	Cardiovascular devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	22/12/2022	22/12/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	22/12/2022	22/12/2022
					CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	22/12/2022	22/12/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	22/12/2022	22/12/2022
					CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	22/12/2022	22/12/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	22/12/2022	22/12/2022
					CT233	Surgical	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	22/12/2022	22/12/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	22/12/2022	22/12/2022
					CT2406	Implantable	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	22/12/2022	22/12/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	22/12/2022	22/12/2022
					CT2407	Active implantable	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	22/12/2022	22/12/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	22/12/2022	22/12/2022
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	22/12/2022	22/12/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	22/12/2022	22/12/2022
					CT2727	Active implant devices/ancillaries	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	22/12/2022	22/12/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	22/12/2022	22/12/2022
					CT2750	Implantable cardiac electrical stimulators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	22/12/2022	22/12/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	22/12/2022	22/12/2022
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	22/12/2022	22/12/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	22/12/2022	22/12/2022
					CT2985	Active	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	22/12/2022	22/12/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	22/12/2022	22/12/2022
					CT3001	Implantable metal	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	22/12/2022	22/12/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	22/12/2022	22/12/2022
					CT3002	Implantable synthetic polymer	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	22/12/2022	22/12/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	22/12/2022	22/12/2022
					CT3006	Implantable inorganic material	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	22/12/2022	22/12/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	22/12/2022	22/12/2022
					CT314	Electrically powered	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	22/12/2022	22/12/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	22/12/2022	22/12/2022
					CT321	Long-term surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	22/12/2022	22/12/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	22/12/2022	22/12/2022
					CT449	Generators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	22/12/2022	22/12/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	22/12/2022	22/12/2022
					CT981	Single use	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	22/12/2022	22/12/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	22/12/2022	22/12/2022
					CT983	Surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	22/12/2022	22/12/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	22/12/2022	22/12/2022
12115	2023		47265	Dual-chamber implantable pacemaker, rate-responsive	CT141	Electrodes/Leads and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	19/12/2022	19/12/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	19/12/2022	19/12/2022
					CT145	Cardiovascular devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	19/12/2022	19/12/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	19/12/2022	19/12/2022
					CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	19/12/2022	19/12/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	19/12/2022	19/12/2022
					CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	19/12/2022	19/12/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	19/12/2022	19/12/2022
					CT233	Surgical	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	19/12/2022	19/12/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	19/12/2022	19/12/2022
					CT2406	Implantable	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	19/12/2022	19/12/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	19/12/2022	19/12/2022
					CT2407	Active implantable	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	19/12/2022	19/12/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	19/12/2022	19/12/2022
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	19/12/2022	19/12/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	19/12/2022	19/12/2022
					CT2727	Active implant devices/ancillaries	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	19/12/2022	19/12/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	19/12/2022	19/12/2022
					CT2750	Implantable cardiac electrical stimulators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	19/12/2022	19/12/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	19/12/2022	19/12/2022
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	19/12/2022	19/12/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	19/12/2022	19/12/2022
					CT2985	Active	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	19/12/2022	19/12/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	19/12/2022	19/12/2022
					CT3001	Implantable metal	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	19/12/2022	19/12/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	19/12/2022	19/12/2022
					CT3002	Implantable synthetic polymer	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	19/12/2022	19/12/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	19/12/2022	19/12/2022
					CT3006	Implantable inorganic material	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	19/12/2022	19/12/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	19/12/2022	19/12/2022
					CT314	Electrically powered	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	19/12/2022	19/12/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	19/12/2022	19/12/2022
					CT321	Long-term surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	19/12/2022	19/12/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	19/12/2022	19/12/2022
					CT449	Generators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	19/12/2022	19/12/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	19/12/2022	19/12/2022
					CT981	Single use	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	19/12/2022	19/12/2022

						F1901	Additional Surgery	A24	No Apparent Device Problem	19/12/2022	19/12/2022
					CT983	Surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	19/12/2022	19/12/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	19/12/2022	19/12/2022
12116	2023		47265	Dual-chamber implantable pacemaker, rate-responsive	CT141	Electrodes/Leads and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	14/01/2023	14/01/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	14/01/2023	14/01/2023
					CT145	Cardiovascular devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	14/01/2023	14/01/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	14/01/2023	14/01/2023
					CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	14/01/2023	14/01/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	14/01/2023	14/01/2023
					CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	14/01/2023	14/01/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	14/01/2023	14/01/2023
					CT233	Surgical	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	14/01/2023	14/01/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	14/01/2023	14/01/2023
					CT2406	Implantable	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	14/01/2023	14/01/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	14/01/2023	14/01/2023
					CT2407	Active implantable	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	14/01/2023	14/01/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	14/01/2023	14/01/2023
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	14/01/2023	14/01/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	14/01/2023	14/01/2023
					CT2727	Active implant devices/ancillaries	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	14/01/2023	14/01/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	14/01/2023	14/01/2023
					CT2750	Implantable cardiac electrical stimulators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	14/01/2023	14/01/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	14/01/2023	14/01/2023
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	14/01/2023	14/01/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	14/01/2023	14/01/2023
					CT2985	Active	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	14/01/2023	14/01/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	14/01/2023	14/01/2023
					CT3001	Implantable metal	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	14/01/2023	14/01/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	14/01/2023	14/01/2023
					CT3002	Implantable synthetic polymer	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	14/01/2023	14/01/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	14/01/2023	14/01/2023
					CT3006	Implantable inorganic material	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	14/01/2023	14/01/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	14/01/2023	14/01/2023
					CT314	Electrically powered	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	14/01/2023	14/01/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	14/01/2023	14/01/2023
					CT321	Long-term surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	14/01/2023	14/01/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	14/01/2023	14/01/2023
					CT449	Generators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	14/01/2023	14/01/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	14/01/2023	14/01/2023
					CT981	Single use	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	14/01/2023	14/01/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	14/01/2023	14/01/2023
					CT983	Surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	14/01/2023	14/01/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	14/01/2023	14/01/2023
12117	2023		47265	Dual-chamber implantable pacemaker, rate-responsive	CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	13/12/2022	13/12/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	13/12/2022	13/12/2022
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	13/12/2022	13/12/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	13/12/2022	13/12/2022
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	13/12/2022	13/12/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	13/12/2022	13/12/2022
12118	2023		47265	Dual-chamber implantable pacemaker, rate-responsive	CT141	Electrodes/Leads and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	01/10/2022	01/10/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	01/10/2022	01/10/2022
					CT145	Cardiovascular devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	01/10/2022	01/10/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	01/10/2022	01/10/2022
					CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	01/10/2022	01/10/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	01/10/2022	01/10/2022
					CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	01/10/2022	01/10/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	01/10/2022	01/10/2022
					CT233	Surgical	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	01/10/2022	01/10/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	01/10/2022	01/10/2022
					CT2406	Implantable	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	01/10/2022	01/10/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	01/10/2022	01/10/2022
					CT2407	Active implantable	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	01/10/2022	01/10/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	01/10/2022	01/10/2022
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	01/10/2022	01/10/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	01/10/2022	01/10/2022
					CT2727	Active implant devices/ancillaries	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	01/10/2022	01/10/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	01/10/2022	01/10/2022
					CT2750	Implantable cardiac electrical stimulators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	01/10/2022	01/10/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	01/10/2022	01/10/2022
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	01/10/2022	01/10/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	01/10/2022	01/10/2022
					CT2985	Active	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	01/10/2022	01/10/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	01/10/2022	01/10/2022
					CT3001	Implantable metal	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	01/10/2022	01/10/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	01/10/2022	01/10/2022
					CT3002	Implantable synthetic polymer	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	01/10/2022	01/10/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	01/10/2022	01/10/2022
					CT3006	Implantable inorganic material	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	01/10/2022	01/10/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	01/10/2022	01/10/2022
					CT314	Electrically powered	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	01/10/2022	01/10/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	01/10/2022	01/10/2022
					CT321	Long-term surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	01/10/2022	01/10/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	01/10/2022	01/10/2022
					CT449	Generators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	01/10/2022	01/10/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	01/10/2022	01/10/2022
					CT981	Single use	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	01/10/2022	01/10/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	01/10/2022	01/10/2022
					CT983	Surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	01/10/2022	01/10/2022

							F1901	Additional Surgery	A24	No Apparent Device Problem	01/10/2022	01/10/2022
12119	2023	All other reportable incidents	47265	Dual-chamber implantable pacemaker, rate-responsive	CT141	Electrodes/Leads and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070101	Failure to Capture		
							F26	No Health Consequences or Impact	A070101	Failure to Capture		
					CT145	Cardiovascular devices	E2403	No Clinical Signs, Symptoms or Conditions	A070101	Failure to Capture		
							F26	No Health Consequences or Impact	A070101	Failure to Capture		
					CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070101	Failure to Capture		
							F26	No Health Consequences or Impact	A070101	Failure to Capture		
					CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A070101	Failure to Capture		
							F26	No Health Consequences or Impact	A070101	Failure to Capture		
					CT233	Surgical	E2403	No Clinical Signs, Symptoms or Conditions	A070101	Failure to Capture		
							F26	No Health Consequences or Impact	A070101	Failure to Capture		
					CT2406	Implantable	E2403	No Clinical Signs, Symptoms or Conditions	A070101	Failure to Capture		
							F26	No Health Consequences or Impact	A070101	Failure to Capture		
					CT2407	Active implantable	E2403	No Clinical Signs, Symptoms or Conditions	A070101	Failure to Capture		
							F26	No Health Consequences or Impact	A070101	Failure to Capture		
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A070101	Failure to Capture		
							F26	No Health Consequences or Impact	A070101	Failure to Capture		
					CT2727	Active implant devices/ancillaries	E2403	No Clinical Signs, Symptoms or Conditions	A070101	Failure to Capture		
							F26	No Health Consequences or Impact	A070101	Failure to Capture		
					CT2750	Implantable cardiac electrical stimulators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070101	Failure to Capture		
							F26	No Health Consequences or Impact	A070101	Failure to Capture		
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A070101	Failure to Capture		
							F26	No Health Consequences or Impact	A070101	Failure to Capture		
					CT2985	Active	E2403	No Clinical Signs, Symptoms or Conditions	A070101	Failure to Capture		
							F26	No Health Consequences or Impact	A070101	Failure to Capture		
					CT3001	Implantable metal	E2403	No Clinical Signs, Symptoms or Conditions	A070101	Failure to Capture		
							F26	No Health Consequences or Impact	A070101	Failure to Capture		
					CT3002	Implantable synthetic polymer	E2403	No Clinical Signs, Symptoms or Conditions	A070101	Failure to Capture		
							F26	No Health Consequences or Impact	A070101	Failure to Capture		
					CT3006	Implantable inorganic material	E2403	No Clinical Signs, Symptoms or Conditions	A070101	Failure to Capture		
							F26	No Health Consequences or Impact	A070101	Failure to Capture		
					CT314	Electrically powered	E2403	No Clinical Signs, Symptoms or Conditions	A070101	Failure to Capture		
							F26	No Health Consequences or Impact	A070101	Failure to Capture		
					CT321	Long-term surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A070101	Failure to Capture		
							F26	No Health Consequences or Impact	A070101	Failure to Capture		
					CT449	Generators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070101	Failure to Capture		
							F26	No Health Consequences or Impact	A070101	Failure to Capture		
					CT981	Single use	E2403	No Clinical Signs, Symptoms or Conditions	A070101	Failure to Capture		
							F26	No Health Consequences or Impact	A070101	Failure to Capture		
					CT983	Surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A070101	Failure to Capture		
							F26	No Health Consequences or Impact	A070101	Failure to Capture		
12120	2023	All other reportable incidents	47267	Single-chamber implantable pacemaker, rate-responsive	CT141	Electrodes/Leads and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070909	Oversensing		
									A071204	No Pacing		
							F22	Unexpected Diagnostic Intervention	A070909	Oversensing		
									A071204	No Pacing		
					CT145	Cardiovascular devices	E2403	No Clinical Signs, Symptoms or Conditions	A070909	Oversensing		
									A071204	No Pacing		
							F22	Unexpected Diagnostic Intervention	A070909	Oversensing		
									A071204	No Pacing		
					CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070909	Oversensing		
									A071204	No Pacing		
							F22	Unexpected Diagnostic Intervention	A070909	Oversensing		
									A071204	No Pacing		
					CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A070909	Oversensing		
									A071204	No Pacing		
							F22	Unexpected Diagnostic Intervention	A070909	Oversensing		
									A071204	No Pacing		
					CT233	Surgical	E2403	No Clinical Signs, Symptoms or Conditions	A070909	Oversensing		
									A071204	No Pacing		
							F22	Unexpected Diagnostic Intervention	A070909	Oversensing		
									A071204	No Pacing		
					CT2406	Implantable	E2403	No Clinical Signs, Symptoms or Conditions	A070909	Oversensing		
									A071204	No Pacing		
							F22	Unexpected Diagnostic Intervention	A070909	Oversensing		
									A071204	No Pacing		
					CT2407	Active implantable	E2403	No Clinical Signs, Symptoms or Conditions	A070909	Oversensing		
									A071204	No Pacing		
							F22	Unexpected Diagnostic Intervention	A070909	Oversensing		
									A071204	No Pacing		
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A070909	Oversensing		
									A071204	No Pacing		
							F22	Unexpected Diagnostic Intervention	A070909	Oversensing		
									A071204	No Pacing		
					CT2727	Active implant devices/ancillaries	E2403	No Clinical Signs, Symptoms or Conditions	A070909	Oversensing		
									A071204	No Pacing		
							F22	Unexpected Diagnostic Intervention	A070909	Oversensing		
									A071204	No Pacing		
					CT2750	Implantable cardiac electrical stimulators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070909	Oversensing		
									A071204	No Pacing		
							F22	Unexpected Diagnostic Intervention	A070909	Oversensing		
									A071204	No Pacing		
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A070909	Oversensing		
									A071204	No Pacing		
							F22	Unexpected Diagnostic Intervention	A070909	Oversensing		
									A071204	No Pacing		
					CT2985	Active	E2403	No Clinical Signs, Symptoms or Conditions	A070909	Oversensing		
									A071204	No Pacing		
							F22	Unexpected Diagnostic Intervention	A070909	Oversensing		

									A071204	No Pacing			
					CT3001	Implantable metal	E2403	No Clinical Signs, Symptoms or Conditions	A070909	Oversensing			
									A071204	No Pacing			
							F22	Unexpected Diagnostic Intervention	A070909	Oversensing			
									A071204	No Pacing			
					CT3002	Implantable synthetic polymer	E2403	No Clinical Signs, Symptoms or Conditions	A070909	Oversensing			
									A071204	No Pacing			
							F22	Unexpected Diagnostic Intervention	A070909	Oversensing			
									A071204	No Pacing			
					CT3006	Implantable inorganic material	E2403	No Clinical Signs, Symptoms or Conditions	A070909	Oversensing			
									A071204	No Pacing			
							F22	Unexpected Diagnostic Intervention	A070909	Oversensing			
									A071204	No Pacing			
					CT314	Electrically powered	E2403	No Clinical Signs, Symptoms or Conditions	A070909	Oversensing			
									A071204	No Pacing			
							F22	Unexpected Diagnostic Intervention	A070909	Oversensing			
									A071204	No Pacing			
					CT321	Long-term surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A070909	Oversensing			
									A071204	No Pacing			
							F22	Unexpected Diagnostic Intervention	A070909	Oversensing			
									A071204	No Pacing			
					CT449	Generators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070909	Oversensing			
									A071204	No Pacing			
							F22	Unexpected Diagnostic Intervention	A070909	Oversensing			
									A071204	No Pacing			
					CT981	Single use	E2403	No Clinical Signs, Symptoms or Conditions	A070909	Oversensing			
									A071204	No Pacing			
							F22	Unexpected Diagnostic Intervention	A070909	Oversensing			
									A071204	No Pacing			
					CT983	Surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A070909	Oversensing			
									A071204	No Pacing			
							F22	Unexpected Diagnostic Intervention	A070909	Oversensing			
									A071204	No Pacing			
12121	2023		47265	Dual-chamber implantable pacemaker, rate-responsive	CT141	Electrodes/Leads and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem			
							F1901	Additional Surgery	A24	No Apparent Device Problem			
					CT145	Cardiovascular devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem			
							F1901	Additional Surgery	A24	No Apparent Device Problem			
					CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem			
							F1901	Additional Surgery	A24	No Apparent Device Problem			
					CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem			
							F1901	Additional Surgery	A24	No Apparent Device Problem			
					CT233	Surgical	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem			
							F1901	Additional Surgery	A24	No Apparent Device Problem			
					CT2406	Implantable	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem			
							F1901	Additional Surgery	A24	No Apparent Device Problem			
					CT2407	Active implantable	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem			
							F1901	Additional Surgery	A24	No Apparent Device Problem			
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem			
							F1901	Additional Surgery	A24	No Apparent Device Problem			
					CT2727	Active implant devices/ancillaries	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem			
							F1901	Additional Surgery	A24	No Apparent Device Problem			
					CT2750	Implantable cardiac electrical stimulators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem			
							F1901	Additional Surgery	A24	No Apparent Device Problem			
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem			
							F1901	Additional Surgery	A24	No Apparent Device Problem			
					CT2985	Active	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem			
							F1901	Additional Surgery	A24	No Apparent Device Problem			
					CT3001	Implantable metal	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem			
							F1901	Additional Surgery	A24	No Apparent Device Problem			
					CT3002	Implantable synthetic polymer	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem			
							F1901	Additional Surgery	A24	No Apparent Device Problem			
					CT3006	Implantable inorganic material	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem			
							F1901	Additional Surgery	A24	No Apparent Device Problem			
					CT314	Electrically powered	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem			
							F1901	Additional Surgery	A24	No Apparent Device Problem			
					CT321	Long-term surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem			
							F1901	Additional Surgery	A24	No Apparent Device Problem			
					CT449	Generators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem			
							F1901	Additional Surgery	A24	No Apparent Device Problem			
					CT981	Single use	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem			
							F1901	Additional Surgery	A24	No Apparent Device Problem			
					CT983	Surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem			
							F1901	Additional Surgery	A24	No Apparent Device Problem			
12122	2023		47265	Dual-chamber implantable pacemaker, rate-responsive	CT141	Electrodes/Leads and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	11/01/2023	11/01/2023	
							F1901	Additional Surgery	A24	No Apparent Device Problem	11/01/2023	11/01/2023	
					CT145	Cardiovascular devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	11/01/2023	11/01/2023	
							F1901	Additional Surgery	A24	No Apparent Device Problem	11/01/2023	11/01/2023	
					CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	11/01/2023	11/01/2023	
							F1901	Additional Surgery	A24	No Apparent Device Problem	11/01/2023	11/01/2023	
					CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	11/01/2023	11/01/2023	
							F1901	Additional Surgery	A24	No Apparent Device Problem	11/01/2023	11/01/2023	
					CT233	Surgical	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	11/01/2023	11/01/2023	
							F1901	Additional Surgery	A24	No Apparent Device Problem	11/01/2023	11/01/2023	
					CT2406	Implantable	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	11/01/2023	11/01/2023	
							F1901	Additional Surgery	A24	No Apparent Device Problem	11/01/2023	11/01/2023	
					CT2407	Active implantable	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	11/01/2023	11/01/2023	
							F1901	Additional Surgery	A24	No Apparent Device Problem	11/01/2023	11/01/2023	
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	11/01/2023	11/01/2023	

						F1901	Additional Surgery	A24	No Apparent Device Problem	11/01/2023	11/01/2023
					CT2727	Active implant devices/ancillaries	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	11/01/2023	11/01/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	11/01/2023	11/01/2023
					CT2750	Implantable cardiac electrical stimulators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	11/01/2023	11/01/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	11/01/2023	11/01/2023
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	11/01/2023	11/01/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	11/01/2023	11/01/2023
					CT2985	Active	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	11/01/2023	11/01/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	11/01/2023	11/01/2023
					CT3001	Implantable metal	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	11/01/2023	11/01/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	11/01/2023	11/01/2023
					CT3002	Implantable synthetic polymer	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	11/01/2023	11/01/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	11/01/2023	11/01/2023
					CT3006	Implantable inorganic material	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	11/01/2023	11/01/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	11/01/2023	11/01/2023
					CT314	Electrically powered	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	11/01/2023	11/01/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	11/01/2023	11/01/2023
					CT321	Long-term surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	11/01/2023	11/01/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	11/01/2023	11/01/2023
					CT449	Generators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	11/01/2023	11/01/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	11/01/2023	11/01/2023
					CT981	Single use	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	11/01/2023	11/01/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	11/01/2023	11/01/2023
					CT983	Surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	11/01/2023	11/01/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	11/01/2023	11/01/2023
12123	2023		47265	Dual-chamber implantable pacemaker, rate-responsive	CT141	Electrodes/Leads and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem		
							F1901	Additional Surgery	A24	No Apparent Device Problem		
					CT145	Cardiovascular devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem		
							F1901	Additional Surgery	A24	No Apparent Device Problem		
					CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem		
							F1901	Additional Surgery	A24	No Apparent Device Problem		
					CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem		
							F1901	Additional Surgery	A24	No Apparent Device Problem		
					CT233	Surgical	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem		
							F1901	Additional Surgery	A24	No Apparent Device Problem		
					CT2406	Implantable	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem		
							F1901	Additional Surgery	A24	No Apparent Device Problem		
					CT2407	Active implantable	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem		
							F1901	Additional Surgery	A24	No Apparent Device Problem		
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem		
							F1901	Additional Surgery	A24	No Apparent Device Problem		
					CT2727	Active implant devices/ancillaries	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem		
							F1901	Additional Surgery	A24	No Apparent Device Problem		
					CT2750	Implantable cardiac electrical stimulators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem		
							F1901	Additional Surgery	A24	No Apparent Device Problem		
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem		
							F1901	Additional Surgery	A24	No Apparent Device Problem		
					CT2985	Active	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem		
							F1901	Additional Surgery	A24	No Apparent Device Problem		
					CT3001	Implantable metal	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem		
							F1901	Additional Surgery	A24	No Apparent Device Problem		
					CT3002	Implantable synthetic polymer	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem		
							F1901	Additional Surgery	A24	No Apparent Device Problem		
					CT3006	Implantable inorganic material	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem		
							F1901	Additional Surgery	A24	No Apparent Device Problem		
					CT314	Electrically powered	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem		
							F1901	Additional Surgery	A24	No Apparent Device Problem		
					CT321	Long-term surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem		
							F1901	Additional Surgery	A24	No Apparent Device Problem		
					CT449	Generators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem		
							F1901	Additional Surgery	A24	No Apparent Device Problem		
					CT981	Single use	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem		
							F1901	Additional Surgery	A24	No Apparent Device Problem		
					CT983	Surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem		
							F1901	Additional Surgery	A24	No Apparent Device Problem		
12124	2023	All other reportable incidents	35852	Single-chamber implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	26/10/2022	26/10/2022
							F1901	Additional Surgery	A070504	Premature Discharge of Battery	26/10/2022	26/10/2022
12125	2023		47265	Dual-chamber implantable pacemaker, rate-responsive	CT141	Electrodes/Leads and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	12/12/2022	12/12/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	12/12/2022	12/12/2022
					CT145	Cardiovascular devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	12/12/2022	12/12/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	12/12/2022	12/12/2022
					CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	12/12/2022	12/12/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	12/12/2022	12/12/2022
					CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	12/12/2022	12/12/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	12/12/2022	12/12/2022
					CT233	Surgical	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	12/12/2022	12/12/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	12/12/2022	12/12/2022
					CT2406	Implantable	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	12/12/2022	12/12/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	12/12/2022	12/12/2022
					CT2407	Active implantable	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	12/12/2022	12/12/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	12/12/2022	12/12/2022
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	12/12/2022	12/12/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	12/12/2022	12/12/2022
					CT2727	Active implant devices/ancillaries	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	12/12/2022	12/12/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	12/12/2022	12/12/2022
					CT2750	Implantable cardiac electrical stimulators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	12/12/2022	12/12/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	12/12/2022	12/12/2022
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	12/12/2022	12/12/2022

							F1901	Additional Surgery	A24	No Apparent Device Problem	12/12/2022	12/12/2022
					CT2985	Active	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	12/12/2022	12/12/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	12/12/2022	12/12/2022
					CT3001	Implantable metal	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	12/12/2022	12/12/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	12/12/2022	12/12/2022
					CT3002	Implantable synthetic polymer	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	12/12/2022	12/12/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	12/12/2022	12/12/2022
					CT3006	Implantable inorganic material	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	12/12/2022	12/12/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	12/12/2022	12/12/2022
					CT314	Electrically powered	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	12/12/2022	12/12/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	12/12/2022	12/12/2022
					CT321	Long-term surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	12/12/2022	12/12/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	12/12/2022	12/12/2022
					CT449	Generators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	12/12/2022	12/12/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	12/12/2022	12/12/2022
					CT981	Single use	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	12/12/2022	12/12/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	12/12/2022	12/12/2022
					CT983	Surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	12/12/2022	12/12/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	12/12/2022	12/12/2022
12126	2023		47265	Dual-chamber implantable pacemaker, rate-responsive	CT141	Electrodes/Leads and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	21/12/2022	21/12/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	21/12/2022	21/12/2022
					CT145	Cardiovascular devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	21/12/2022	21/12/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	21/12/2022	21/12/2022
					CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	21/12/2022	21/12/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	21/12/2022	21/12/2022
					CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	21/12/2022	21/12/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	21/12/2022	21/12/2022
					CT233	Surgical	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	21/12/2022	21/12/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	21/12/2022	21/12/2022
					CT2406	Implantable	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	21/12/2022	21/12/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	21/12/2022	21/12/2022
					CT2407	Active implantable	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	21/12/2022	21/12/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	21/12/2022	21/12/2022
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	21/12/2022	21/12/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	21/12/2022	21/12/2022
					CT2727	Active implant devices/ancillaries	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	21/12/2022	21/12/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	21/12/2022	21/12/2022
					CT2750	Implantable cardiac electrical stimulators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	21/12/2022	21/12/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	21/12/2022	21/12/2022
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	21/12/2022	21/12/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	21/12/2022	21/12/2022
					CT2985	Active	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	21/12/2022	21/12/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	21/12/2022	21/12/2022
					CT3001	Implantable metal	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	21/12/2022	21/12/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	21/12/2022	21/12/2022
					CT3002	Implantable synthetic polymer	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	21/12/2022	21/12/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	21/12/2022	21/12/2022
					CT3006	Implantable inorganic material	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	21/12/2022	21/12/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	21/12/2022	21/12/2022
					CT314	Electrically powered	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	21/12/2022	21/12/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	21/12/2022	21/12/2022
					CT321	Long-term surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	21/12/2022	21/12/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	21/12/2022	21/12/2022
					CT449	Generators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	21/12/2022	21/12/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	21/12/2022	21/12/2022
					CT981	Single use	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	21/12/2022	21/12/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	21/12/2022	21/12/2022
					CT983	Surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	21/12/2022	21/12/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	21/12/2022	21/12/2022
12127	2023		47265	Dual-chamber implantable pacemaker, rate-responsive	CT141	Electrodes/Leads and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	18/01/2023	18/01/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	18/01/2023	18/01/2023
					CT145	Cardiovascular devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	18/01/2023	18/01/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	18/01/2023	18/01/2023
					CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	18/01/2023	18/01/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	18/01/2023	18/01/2023
					CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	18/01/2023	18/01/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	18/01/2023	18/01/2023
					CT233	Surgical	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	18/01/2023	18/01/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	18/01/2023	18/01/2023
					CT2406	Implantable	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	18/01/2023	18/01/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	18/01/2023	18/01/2023
					CT2407	Active implantable	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	18/01/2023	18/01/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	18/01/2023	18/01/2023
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	18/01/2023	18/01/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	18/01/2023	18/01/2023
					CT2727	Active implant devices/ancillaries	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	18/01/2023	18/01/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	18/01/2023	18/01/2023
					CT2750	Implantable cardiac electrical stimulators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	18/01/2023	18/01/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	18/01/2023	18/01/2023
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	18/01/2023	18/01/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	18/01/2023	18/01/2023
					CT2985	Active	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	18/01/2023	18/01/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	18/01/2023	18/01/2023
					CT3001	Implantable metal	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	18/01/2023	18/01/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	18/01/2023	18/01/2023
					CT3002	Implantable synthetic polymer	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	18/01/2023	18/01/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	18/01/2023	18/01/2023
					CT3006	Implantable inorganic material	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	18/01/2023	18/01/2023

						F1901	Additional Surgery	A24	No Apparent Device Problem	18/01/2023	18/01/2023
					CT314	Electrically powered	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	18/01/2023	18/01/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	18/01/2023	18/01/2023
					CT321	Long-term surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	18/01/2023	18/01/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	18/01/2023	18/01/2023
					CT449	Generators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	18/01/2023	18/01/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	18/01/2023	18/01/2023
					CT981	Single use	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	18/01/2023	18/01/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	18/01/2023	18/01/2023
					CT983	Surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	18/01/2023	18/01/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	18/01/2023	18/01/2023
12128	2023		47265	Dual-chamber implantable pacemaker, rate-responsive	CT141	Electrodes/Leads and associated devices	E0601	Arrhythmia	A071204	No Pacing	02/11/2022	02/11/2022
							F1908	Prolonged Surgery	A071204	No Pacing	02/11/2022	02/11/2022
					CT145	Cardiovascular devices	E0601	Arrhythmia	A071204	No Pacing	02/11/2022	02/11/2022
							F1908	Prolonged Surgery	A071204	No Pacing	02/11/2022	02/11/2022
					CT1684	Cardiac pacemakers and associated devices	E0601	Arrhythmia	A071204	No Pacing	02/11/2022	02/11/2022
							F1908	Prolonged Surgery	A071204	No Pacing	02/11/2022	02/11/2022
					CT1685	Cardiac pacemakers	E0601	Arrhythmia	A071204	No Pacing	02/11/2022	02/11/2022
							F1908	Prolonged Surgery	A071204	No Pacing	02/11/2022	02/11/2022
					CT233	Surgical	E0601	Arrhythmia	A071204	No Pacing	02/11/2022	02/11/2022
							F1908	Prolonged Surgery	A071204	No Pacing	02/11/2022	02/11/2022
					CT2406	Implantable	E0601	Arrhythmia	A071204	No Pacing	02/11/2022	02/11/2022
							F1908	Prolonged Surgery	A071204	No Pacing	02/11/2022	02/11/2022
					CT2407	Active implantable	E0601	Arrhythmia	A071204	No Pacing	02/11/2022	02/11/2022
							F1908	Prolonged Surgery	A071204	No Pacing	02/11/2022	02/11/2022
					CT257	Implantable cardiac pacemakers	E0601	Arrhythmia	A071204	No Pacing	02/11/2022	02/11/2022
							F1908	Prolonged Surgery	A071204	No Pacing	02/11/2022	02/11/2022
					CT2727	Active implant devices/ancillaries	E0601	Arrhythmia	A071204	No Pacing	02/11/2022	02/11/2022
							F1908	Prolonged Surgery	A071204	No Pacing	02/11/2022	02/11/2022
					CT2750	Implantable cardiac electrical stimulators and associated devices	E0601	Arrhythmia	A071204	No Pacing	02/11/2022	02/11/2022
							F1908	Prolonged Surgery	A071204	No Pacing	02/11/2022	02/11/2022
					CT2866	Implantable cardiac electrical stimulators	E0601	Arrhythmia	A071204	No Pacing	02/11/2022	02/11/2022
							F1908	Prolonged Surgery	A071204	No Pacing	02/11/2022	02/11/2022
					CT2985	Active	E0601	Arrhythmia	A071204	No Pacing	02/11/2022	02/11/2022
							F1908	Prolonged Surgery	A071204	No Pacing	02/11/2022	02/11/2022
					CT3001	Implantable metal	E0601	Arrhythmia	A071204	No Pacing	02/11/2022	02/11/2022
							F1908	Prolonged Surgery	A071204	No Pacing	02/11/2022	02/11/2022
					CT3002	Implantable synthetic polymer	E0601	Arrhythmia	A071204	No Pacing	02/11/2022	02/11/2022
							F1908	Prolonged Surgery	A071204	No Pacing	02/11/2022	02/11/2022
					CT3006	Implantable inorganic material	E0601	Arrhythmia	A071204	No Pacing	02/11/2022	02/11/2022
							F1908	Prolonged Surgery	A071204	No Pacing	02/11/2022	02/11/2022
					CT314	Electrically powered	E0601	Arrhythmia	A071204	No Pacing	02/11/2022	02/11/2022
							F1908	Prolonged Surgery	A071204	No Pacing	02/11/2022	02/11/2022
					CT321	Long-term surgical invasive	E0601	Arrhythmia	A071204	No Pacing	02/11/2022	02/11/2022
							F1908	Prolonged Surgery	A071204	No Pacing	02/11/2022	02/11/2022
					CT449	Generators and associated devices	E0601	Arrhythmia	A071204	No Pacing	02/11/2022	02/11/2022
							F1908	Prolonged Surgery	A071204	No Pacing	02/11/2022	02/11/2022
					CT981	Single use	E0601	Arrhythmia	A071204	No Pacing	02/11/2022	02/11/2022
							F1908	Prolonged Surgery	A071204	No Pacing	02/11/2022	02/11/2022
					CT983	Surgical invasive	E0601	Arrhythmia	A071204	No Pacing	02/11/2022	02/11/2022
							F1908	Prolonged Surgery	A071204	No Pacing	02/11/2022	02/11/2022
12129	2023		47265	Dual-chamber implantable pacemaker, rate-responsive	CT141	Electrodes/Leads and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	01/12/2022	01/12/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	01/12/2022	01/12/2022
					CT145	Cardiovascular devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	01/12/2022	01/12/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	01/12/2022	01/12/2022
					CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	01/12/2022	01/12/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	01/12/2022	01/12/2022
					CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	01/12/2022	01/12/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	01/12/2022	01/12/2022
					CT233	Surgical	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	01/12/2022	01/12/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	01/12/2022	01/12/2022
					CT2406	Implantable	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	01/12/2022	01/12/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	01/12/2022	01/12/2022
					CT2407	Active implantable	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	01/12/2022	01/12/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	01/12/2022	01/12/2022
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	01/12/2022	01/12/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	01/12/2022	01/12/2022
					CT2727	Active implant devices/ancillaries	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	01/12/2022	01/12/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	01/12/2022	01/12/2022
					CT2750	Implantable cardiac electrical stimulators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	01/12/2022	01/12/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	01/12/2022	01/12/2022
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	01/12/2022	01/12/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	01/12/2022	01/12/2022
					CT2985	Active	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	01/12/2022	01/12/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	01/12/2022	01/12/2022
					CT3001	Implantable metal	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	01/12/2022	01/12/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	01/12/2022	01/12/2022
					CT3002	Implantable synthetic polymer	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	01/12/2022	01/12/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	01/12/2022	01/12/2022
					CT3006	Implantable inorganic material	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	01/12/2022	01/12/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	01/12/2022	01/12/2022
					CT314	Electrically powered	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	01/12/2022	01/12/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	01/12/2022	01/12/2022
					CT321	Long-term surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	01/12/2022	01/12/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	01/12/2022	01/12/2022
					CT449	Generators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	01/12/2022	01/12/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	01/12/2022	01/12/2022
					CT981	Single use	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	01/12/2022	01/12/2022

						F1901	Additional Surgery	A24	No Apparent Device Problem	01/12/2022	01/12/2022
					CT983	Surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	01/12/2022	01/12/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	01/12/2022	01/12/2022
12130	2023		47265	Dual-chamber implantable pacemaker, rate-responsive	CT141	Electrodes/Leads and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	21/12/2022	21/12/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	21/12/2022	21/12/2022
					CT145	Cardiovascular devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	21/12/2022	21/12/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	21/12/2022	21/12/2022
					CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	21/12/2022	21/12/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	21/12/2022	21/12/2022
					CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	21/12/2022	21/12/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	21/12/2022	21/12/2022
					CT233	Surgical	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	21/12/2022	21/12/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	21/12/2022	21/12/2022
					CT2406	Implantable	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	21/12/2022	21/12/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	21/12/2022	21/12/2022
					CT2407	Active implantable	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	21/12/2022	21/12/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	21/12/2022	21/12/2022
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	21/12/2022	21/12/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	21/12/2022	21/12/2022
					CT2727	Active implant devices/ancillaries	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	21/12/2022	21/12/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	21/12/2022	21/12/2022
					CT2750	Implantable cardiac electrical stimulators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	21/12/2022	21/12/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	21/12/2022	21/12/2022
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	21/12/2022	21/12/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	21/12/2022	21/12/2022
					CT2985	Active	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	21/12/2022	21/12/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	21/12/2022	21/12/2022
					CT3001	Implantable metal	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	21/12/2022	21/12/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	21/12/2022	21/12/2022
					CT3002	Implantable synthetic polymer	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	21/12/2022	21/12/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	21/12/2022	21/12/2022
					CT3006	Implantable inorganic material	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	21/12/2022	21/12/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	21/12/2022	21/12/2022
					CT314	Electrically powered	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	21/12/2022	21/12/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	21/12/2022	21/12/2022
					CT321	Long-term surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	21/12/2022	21/12/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	21/12/2022	21/12/2022
					CT449	Generators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	21/12/2022	21/12/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	21/12/2022	21/12/2022
					CT981	Single use	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	21/12/2022	21/12/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	21/12/2022	21/12/2022
					CT983	Surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	21/12/2022	21/12/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	21/12/2022	21/12/2022
12131	2023		47265	Dual-chamber implantable pacemaker, rate-responsive	CT141	Electrodes/Leads and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	18/01/2023	18/01/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	18/01/2023	18/01/2023
					CT145	Cardiovascular devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	18/01/2023	18/01/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	18/01/2023	18/01/2023
					CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	18/01/2023	18/01/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	18/01/2023	18/01/2023
					CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	18/01/2023	18/01/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	18/01/2023	18/01/2023
					CT233	Surgical	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	18/01/2023	18/01/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	18/01/2023	18/01/2023
					CT2406	Implantable	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	18/01/2023	18/01/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	18/01/2023	18/01/2023
					CT2407	Active implantable	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	18/01/2023	18/01/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	18/01/2023	18/01/2023
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	18/01/2023	18/01/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	18/01/2023	18/01/2023
					CT2727	Active implant devices/ancillaries	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	18/01/2023	18/01/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	18/01/2023	18/01/2023
					CT2750	Implantable cardiac electrical stimulators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	18/01/2023	18/01/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	18/01/2023	18/01/2023
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	18/01/2023	18/01/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	18/01/2023	18/01/2023
					CT2985	Active	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	18/01/2023	18/01/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	18/01/2023	18/01/2023
					CT3001	Implantable metal	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	18/01/2023	18/01/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	18/01/2023	18/01/2023
					CT3002	Implantable synthetic polymer	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	18/01/2023	18/01/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	18/01/2023	18/01/2023
					CT3006	Implantable inorganic material	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	18/01/2023	18/01/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	18/01/2023	18/01/2023
					CT314	Electrically powered	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	18/01/2023	18/01/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	18/01/2023	18/01/2023
					CT321	Long-term surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	18/01/2023	18/01/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	18/01/2023	18/01/2023
					CT449	Generators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	18/01/2023	18/01/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	18/01/2023	18/01/2023
					CT981	Single use	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	18/01/2023	18/01/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	18/01/2023	18/01/2023
					CT983	Surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	18/01/2023	18/01/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	18/01/2023	18/01/2023
12132	2023		47265	Dual-chamber implantable pacemaker, rate-responsive	CT141	Electrodes/Leads and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	14/12/2022	14/12/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	14/12/2022	14/12/2022
					CT145	Cardiovascular devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	14/12/2022	14/12/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	14/12/2022	14/12/2022
					CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	14/12/2022	14/12/2022

							F1901	Additional Surgery	A24	No Apparent Device Problem	14/12/2022	14/12/2022
					CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	14/12/2022	14/12/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	14/12/2022	14/12/2022
					CT233	Surgical	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	14/12/2022	14/12/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	14/12/2022	14/12/2022
					CT2406	Implantable	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	14/12/2022	14/12/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	14/12/2022	14/12/2022
					CT2407	Active implantable	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	14/12/2022	14/12/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	14/12/2022	14/12/2022
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	14/12/2022	14/12/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	14/12/2022	14/12/2022
					CT2727	Active implant devices/ancillaries	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	14/12/2022	14/12/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	14/12/2022	14/12/2022
					CT2750	Implantable cardiac electrical stimulators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	14/12/2022	14/12/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	14/12/2022	14/12/2022
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	14/12/2022	14/12/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	14/12/2022	14/12/2022
					CT2985	Active	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	14/12/2022	14/12/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	14/12/2022	14/12/2022
					CT3001	Implantable metal	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	14/12/2022	14/12/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	14/12/2022	14/12/2022
					CT3002	Implantable synthetic polymer	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	14/12/2022	14/12/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	14/12/2022	14/12/2022
					CT3006	Implantable inorganic material	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	14/12/2022	14/12/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	14/12/2022	14/12/2022
					CT314	Electrically powered	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	14/12/2022	14/12/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	14/12/2022	14/12/2022
					CT321	Long-term surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	14/12/2022	14/12/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	14/12/2022	14/12/2022
					CT449	Generators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	14/12/2022	14/12/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	14/12/2022	14/12/2022
					CT981	Single use	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	14/12/2022	14/12/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	14/12/2022	14/12/2022
					CT983	Surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	14/12/2022	14/12/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	14/12/2022	14/12/2022
12133	2023	All other reportable incidents	47265	Dual-chamber implantable pacemaker, rate-responsive	CT141	Electrodes/Leads and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem		
							F1901	Additional Surgery	A24	No Apparent Device Problem		
					CT145	Cardiovascular devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem		
							F1901	Additional Surgery	A24	No Apparent Device Problem		
					CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem		
							F1901	Additional Surgery	A24	No Apparent Device Problem		
					CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem		
							F1901	Additional Surgery	A24	No Apparent Device Problem		
					CT233	Surgical	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem		
							F1901	Additional Surgery	A24	No Apparent Device Problem		
					CT2406	Implantable	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem		
							F1901	Additional Surgery	A24	No Apparent Device Problem		
					CT2407	Active implantable	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem		
							F1901	Additional Surgery	A24	No Apparent Device Problem		
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem		
							F1901	Additional Surgery	A24	No Apparent Device Problem		
					CT2727	Active implant devices/ancillaries	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem		
							F1901	Additional Surgery	A24	No Apparent Device Problem		
					CT2750	Implantable cardiac electrical stimulators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem		
							F1901	Additional Surgery	A24	No Apparent Device Problem		
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem		
							F1901	Additional Surgery	A24	No Apparent Device Problem		
					CT2985	Active	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem		
							F1901	Additional Surgery	A24	No Apparent Device Problem		
					CT3001	Implantable metal	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem		
							F1901	Additional Surgery	A24	No Apparent Device Problem		
					CT3002	Implantable synthetic polymer	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem		
							F1901	Additional Surgery	A24	No Apparent Device Problem		
					CT3006	Implantable inorganic material	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem		
							F1901	Additional Surgery	A24	No Apparent Device Problem		
					CT314	Electrically powered	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem		
							F1901	Additional Surgery	A24	No Apparent Device Problem		
					CT321	Long-term surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem		
							F1901	Additional Surgery	A24	No Apparent Device Problem		
					CT449	Generators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem		
							F1901	Additional Surgery	A24	No Apparent Device Problem		
					CT981	Single use	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem		
							F1901	Additional Surgery	A24	No Apparent Device Problem		
					CT983	Surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem		
							F1901	Additional Surgery	A24	No Apparent Device Problem		
12134	2023		47265	Dual-chamber implantable pacemaker, rate-responsive	CT141	Electrodes/Leads and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	09/01/2023	09/01/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	09/01/2023	09/01/2023
					CT145	Cardiovascular devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	09/01/2023	09/01/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	09/01/2023	09/01/2023
					CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	09/01/2023	09/01/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	09/01/2023	09/01/2023
					CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	09/01/2023	09/01/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	09/01/2023	09/01/2023
					CT233	Surgical	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	09/01/2023	09/01/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	09/01/2023	09/01/2023
					CT2406	Implantable	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	09/01/2023	09/01/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	09/01/2023	09/01/2023
					CT2407	Active implantable	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	09/01/2023	09/01/2023

							F1901	Additional Surgery	A24	No Apparent Device Problem	09/01/2023	09/01/2023
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	09/01/2023	09/01/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	09/01/2023	09/01/2023
					CT2727	Active implant devices/ancillaries	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	09/01/2023	09/01/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	09/01/2023	09/01/2023
					CT2750	Implantable cardiac electrical stimulators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	09/01/2023	09/01/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	09/01/2023	09/01/2023
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	09/01/2023	09/01/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	09/01/2023	09/01/2023
					CT2985	Active	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	09/01/2023	09/01/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	09/01/2023	09/01/2023
					CT3001	Implantable metal	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	09/01/2023	09/01/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	09/01/2023	09/01/2023
					CT3002	Implantable synthetic polymer	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	09/01/2023	09/01/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	09/01/2023	09/01/2023
					CT3006	Implantable inorganic material	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	09/01/2023	09/01/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	09/01/2023	09/01/2023
					CT314	Electrically powered	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	09/01/2023	09/01/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	09/01/2023	09/01/2023
					CT321	Long-term surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	09/01/2023	09/01/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	09/01/2023	09/01/2023
					CT449	Generators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	09/01/2023	09/01/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	09/01/2023	09/01/2023
					CT981	Single use	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	09/01/2023	09/01/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	09/01/2023	09/01/2023
					CT983	Surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	09/01/2023	09/01/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	09/01/2023	09/01/2023
12135	2023		47265	Dual-chamber implantable pacemaker, rate-responsive	CT141	Electrodes/Leads and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem		
							F1901	Additional Surgery	A24	No Apparent Device Problem		
					CT145	Cardiovascular devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem		
							F1901	Additional Surgery	A24	No Apparent Device Problem		
					CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem		
							F1901	Additional Surgery	A24	No Apparent Device Problem		
					CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem		
							F1901	Additional Surgery	A24	No Apparent Device Problem		
					CT233	Surgical	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem		
							F1901	Additional Surgery	A24	No Apparent Device Problem		
					CT2406	Implantable	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem		
							F1901	Additional Surgery	A24	No Apparent Device Problem		
					CT2407	Active implantable	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem		
							F1901	Additional Surgery	A24	No Apparent Device Problem		
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem		
							F1901	Additional Surgery	A24	No Apparent Device Problem		
					CT2727	Active implant devices/ancillaries	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem		
							F1901	Additional Surgery	A24	No Apparent Device Problem		
					CT2750	Implantable cardiac electrical stimulators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem		
							F1901	Additional Surgery	A24	No Apparent Device Problem		
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem		
							F1901	Additional Surgery	A24	No Apparent Device Problem		
					CT2985	Active	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem		
							F1901	Additional Surgery	A24	No Apparent Device Problem		
					CT3001	Implantable metal	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem		
							F1901	Additional Surgery	A24	No Apparent Device Problem		
					CT3002	Implantable synthetic polymer	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem		
							F1901	Additional Surgery	A24	No Apparent Device Problem		
					CT3006	Implantable inorganic material	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem		
							F1901	Additional Surgery	A24	No Apparent Device Problem		
					CT314	Electrically powered	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem		
							F1901	Additional Surgery	A24	No Apparent Device Problem		
					CT321	Long-term surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem		
							F1901	Additional Surgery	A24	No Apparent Device Problem		
					CT449	Generators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem		
							F1901	Additional Surgery	A24	No Apparent Device Problem		
					CT981	Single use	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem		
							F1901	Additional Surgery	A24	No Apparent Device Problem		
					CT983	Surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem		
							F1901	Additional Surgery	A24	No Apparent Device Problem		
12136	2023		47265	Dual-chamber implantable pacemaker, rate-responsive	CT141	Electrodes/Leads and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	14/12/2022	14/12/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	14/12/2022	14/12/2022
					CT145	Cardiovascular devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	14/12/2022	14/12/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	14/12/2022	14/12/2022
					CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	14/12/2022	14/12/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	14/12/2022	14/12/2022
					CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	14/12/2022	14/12/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	14/12/2022	14/12/2022
					CT233	Surgical	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	14/12/2022	14/12/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	14/12/2022	14/12/2022
					CT2406	Implantable	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	14/12/2022	14/12/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	14/12/2022	14/12/2022
					CT2407	Active implantable	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	14/12/2022	14/12/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	14/12/2022	14/12/2022
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	14/12/2022	14/12/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	14/12/2022	14/12/2022
					CT2727	Active implant devices/ancillaries	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	14/12/2022	14/12/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	14/12/2022	14/12/2022
					CT2750	Implantable cardiac electrical stimulators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	14/12/2022	14/12/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	14/12/2022	14/12/2022
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	14/12/2022	14/12/2022

							F1901	Additional Surgery	A24	No Apparent Device Problem	14/12/2022	14/12/2022
					CT2985	Active	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	14/12/2022	14/12/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	14/12/2022	14/12/2022
					CT3001	Implantable metal	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	14/12/2022	14/12/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	14/12/2022	14/12/2022
					CT3002	Implantable synthetic polymer	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	14/12/2022	14/12/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	14/12/2022	14/12/2022
					CT3006	Implantable inorganic material	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	14/12/2022	14/12/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	14/12/2022	14/12/2022
					CT314	Electrically powered	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	14/12/2022	14/12/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	14/12/2022	14/12/2022
					CT321	Long-term surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	14/12/2022	14/12/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	14/12/2022	14/12/2022
					CT449	Generators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	14/12/2022	14/12/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	14/12/2022	14/12/2022
					CT981	Single use	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	14/12/2022	14/12/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	14/12/2022	14/12/2022
					CT983	Surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	14/12/2022	14/12/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	14/12/2022	14/12/2022
12137	2023		47265	Dual-chamber implantable pacemaker, rate-responsive	CT141	Electrodes/Leads and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	09/01/2023	09/01/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	09/01/2023	09/01/2023
					CT145	Cardiovascular devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	09/01/2023	09/01/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	09/01/2023	09/01/2023
					CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	09/01/2023	09/01/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	09/01/2023	09/01/2023
					CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	09/01/2023	09/01/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	09/01/2023	09/01/2023
					CT233	Surgical	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	09/01/2023	09/01/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	09/01/2023	09/01/2023
					CT2406	Implantable	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	09/01/2023	09/01/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	09/01/2023	09/01/2023
					CT2407	Active implantable	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	09/01/2023	09/01/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	09/01/2023	09/01/2023
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	09/01/2023	09/01/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	09/01/2023	09/01/2023
					CT2727	Active implant devices/ancillaries	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	09/01/2023	09/01/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	09/01/2023	09/01/2023
					CT2750	Implantable cardiac electrical stimulators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	09/01/2023	09/01/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	09/01/2023	09/01/2023
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	09/01/2023	09/01/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	09/01/2023	09/01/2023
					CT2985	Active	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	09/01/2023	09/01/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	09/01/2023	09/01/2023
					CT3001	Implantable metal	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	09/01/2023	09/01/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	09/01/2023	09/01/2023
					CT3002	Implantable synthetic polymer	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	09/01/2023	09/01/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	09/01/2023	09/01/2023
					CT3006	Implantable inorganic material	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	09/01/2023	09/01/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	09/01/2023	09/01/2023
					CT314	Electrically powered	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	09/01/2023	09/01/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	09/01/2023	09/01/2023
					CT321	Long-term surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	09/01/2023	09/01/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	09/01/2023	09/01/2023
					CT449	Generators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	09/01/2023	09/01/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	09/01/2023	09/01/2023
					CT981	Single use	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	09/01/2023	09/01/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	09/01/2023	09/01/2023
					CT983	Surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	09/01/2023	09/01/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	09/01/2023	09/01/2023
12138	2023	All other reportable incidents	17882	Physiologic-monitoring defibrillation system	CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A071204	No Pacing		
							F27	Problem Identified During Non-Clinical Procedure	A071204	No Pacing		
12139	2023	All other reportable incidents	17882	Physiologic-monitoring defibrillation system	CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A26	Insufficient Device Problem Information		
							F27	Problem Identified During Non-Clinical Procedure	A26	Insufficient Device Problem Information		
12140	2023		47265	Dual-chamber implantable pacemaker, rate-responsive	CT141	Electrodes/Leads and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	06/01/2023	06/01/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	06/01/2023	06/01/2023
					CT145	Cardiovascular devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	06/01/2023	06/01/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	06/01/2023	06/01/2023
					CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	06/01/2023	06/01/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	06/01/2023	06/01/2023
					CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	06/01/2023	06/01/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	06/01/2023	06/01/2023
					CT233	Surgical	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	06/01/2023	06/01/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	06/01/2023	06/01/2023
					CT2406	Implantable	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	06/01/2023	06/01/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	06/01/2023	06/01/2023
					CT2407	Active implantable	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	06/01/2023	06/01/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	06/01/2023	06/01/2023
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	06/01/2023	06/01/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	06/01/2023	06/01/2023
					CT2727	Active implant devices/ancillaries	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	06/01/2023	06/01/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	06/01/2023	06/01/2023
					CT2750	Implantable cardiac electrical stimulators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	06/01/2023	06/01/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	06/01/2023	06/01/2023
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	06/01/2023	06/01/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	06/01/2023	06/01/2023
					CT2985	Active	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	06/01/2023	06/01/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	06/01/2023	06/01/2023
					CT3001	Implantable metal	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	06/01/2023	06/01/2023

						F1901	Additional Surgery	A24	No Apparent Device Problem	06/01/2023	06/01/2023
					CT3002	Implantable synthetic polymer	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	06/01/2023	06/01/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	06/01/2023	06/01/2023
					CT3006	Implantable inorganic material	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	06/01/2023	06/01/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	06/01/2023	06/01/2023
					CT314	Electrically powered	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	06/01/2023	06/01/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	06/01/2023	06/01/2023
					CT321	Long-term surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	06/01/2023	06/01/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	06/01/2023	06/01/2023
					CT449	Generators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	06/01/2023	06/01/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	06/01/2023	06/01/2023
					CT981	Single use	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	06/01/2023	06/01/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	06/01/2023	06/01/2023
					CT983	Surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	06/01/2023	06/01/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	06/01/2023	06/01/2023
12141	2023		47265	Dual-chamber implantable pacemaker, rate-responsive	CT141	Electrodes/Leads and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	18/01/2023	18/01/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	18/01/2023	18/01/2023
					CT145	Cardiovascular devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	18/01/2023	18/01/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	18/01/2023	18/01/2023
					CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	18/01/2023	18/01/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	18/01/2023	18/01/2023
					CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	18/01/2023	18/01/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	18/01/2023	18/01/2023
					CT233	Surgical	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	18/01/2023	18/01/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	18/01/2023	18/01/2023
					CT2406	Implantable	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	18/01/2023	18/01/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	18/01/2023	18/01/2023
					CT2407	Active implantable	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	18/01/2023	18/01/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	18/01/2023	18/01/2023
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	18/01/2023	18/01/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	18/01/2023	18/01/2023
					CT2727	Active implant devices/ancillaries	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	18/01/2023	18/01/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	18/01/2023	18/01/2023
					CT2750	Implantable cardiac electrical stimulators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	18/01/2023	18/01/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	18/01/2023	18/01/2023
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	18/01/2023	18/01/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	18/01/2023	18/01/2023
					CT2985	Active	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	18/01/2023	18/01/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	18/01/2023	18/01/2023
					CT3001	Implantable metal	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	18/01/2023	18/01/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	18/01/2023	18/01/2023
					CT3002	Implantable synthetic polymer	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	18/01/2023	18/01/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	18/01/2023	18/01/2023
					CT3006	Implantable inorganic material	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	18/01/2023	18/01/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	18/01/2023	18/01/2023
					CT314	Electrically powered	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	18/01/2023	18/01/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	18/01/2023	18/01/2023
					CT321	Long-term surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	18/01/2023	18/01/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	18/01/2023	18/01/2023
					CT449	Generators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	18/01/2023	18/01/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	18/01/2023	18/01/2023
					CT981	Single use	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	18/01/2023	18/01/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	18/01/2023	18/01/2023
					CT983	Surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	18/01/2023	18/01/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	18/01/2023	18/01/2023
12142	2023		47265	Dual-chamber implantable pacemaker, rate-responsive	CT141	Electrodes/Leads and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	14/12/2022	14/12/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	14/12/2022	14/12/2022
					CT145	Cardiovascular devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	14/12/2022	14/12/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	14/12/2022	14/12/2022
					CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	14/12/2022	14/12/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	14/12/2022	14/12/2022
					CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	14/12/2022	14/12/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	14/12/2022	14/12/2022
					CT233	Surgical	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	14/12/2022	14/12/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	14/12/2022	14/12/2022
					CT2406	Implantable	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	14/12/2022	14/12/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	14/12/2022	14/12/2022
					CT2407	Active implantable	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	14/12/2022	14/12/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	14/12/2022	14/12/2022
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	14/12/2022	14/12/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	14/12/2022	14/12/2022
					CT2727	Active implant devices/ancillaries	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	14/12/2022	14/12/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	14/12/2022	14/12/2022
					CT2750	Implantable cardiac electrical stimulators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	14/12/2022	14/12/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	14/12/2022	14/12/2022
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	14/12/2022	14/12/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	14/12/2022	14/12/2022
					CT2985	Active	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	14/12/2022	14/12/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	14/12/2022	14/12/2022
					CT3001	Implantable metal	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	14/12/2022	14/12/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	14/12/2022	14/12/2022
					CT3002	Implantable synthetic polymer	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	14/12/2022	14/12/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	14/12/2022	14/12/2022
					CT3006	Implantable inorganic material	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	14/12/2022	14/12/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	14/12/2022	14/12/2022
					CT314	Electrically powered	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	14/12/2022	14/12/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	14/12/2022	14/12/2022
					CT321	Long-term surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	14/12/2022	14/12/2022

						F1901	Additional Surgery	A24	No Apparent Device Problem	14/12/2022	14/12/2022
				CT449	Generators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	14/12/2022	14/12/2022
						F1901	Additional Surgery	A24	No Apparent Device Problem	14/12/2022	14/12/2022
				CT981	Single use	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	14/12/2022	14/12/2022
						F1901	Additional Surgery	A24	No Apparent Device Problem	14/12/2022	14/12/2022
				CT983	Surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	14/12/2022	14/12/2022
						F1901	Additional Surgery	A24	No Apparent Device Problem	14/12/2022	14/12/2022
12143	2023		47265	Dual-chamber implantable pacemaker, rate-responsive	CT141	Electrodes/Leads and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	13/12/2022	13/12/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	13/12/2022	13/12/2022
					CT145	Cardiovascular devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	13/12/2022	13/12/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	13/12/2022	13/12/2022
					CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	13/12/2022	13/12/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	13/12/2022	13/12/2022
					CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	13/12/2022	13/12/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	13/12/2022	13/12/2022
					CT233	Surgical	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	13/12/2022	13/12/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	13/12/2022	13/12/2022
					CT2406	Implantable	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	13/12/2022	13/12/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	13/12/2022	13/12/2022
					CT2407	Active implantable	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	13/12/2022	13/12/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	13/12/2022	13/12/2022
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	13/12/2022	13/12/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	13/12/2022	13/12/2022
					CT2727	Active implant devices/ancillaries	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	13/12/2022	13/12/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	13/12/2022	13/12/2022
					CT2750	Implantable cardiac electrical stimulators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	13/12/2022	13/12/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	13/12/2022	13/12/2022
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	13/12/2022	13/12/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	13/12/2022	13/12/2022
					CT2985	Active	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	13/12/2022	13/12/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	13/12/2022	13/12/2022
					CT3001	Implantable metal	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	13/12/2022	13/12/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	13/12/2022	13/12/2022
					CT3002	Implantable synthetic polymer	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	13/12/2022	13/12/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	13/12/2022	13/12/2022
					CT3006	Implantable inorganic material	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	13/12/2022	13/12/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	13/12/2022	13/12/2022
					CT314	Electrically powered	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	13/12/2022	13/12/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	13/12/2022	13/12/2022
					CT321	Long-term surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	13/12/2022	13/12/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	13/12/2022	13/12/2022
					CT449	Generators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	13/12/2022	13/12/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	13/12/2022	13/12/2022
					CT981	Single use	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	13/12/2022	13/12/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	13/12/2022	13/12/2022
					CT983	Surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	13/12/2022	13/12/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	13/12/2022	13/12/2022
12144	2023		47265	Dual-chamber implantable pacemaker, rate-responsive	CT141	Electrodes/Leads and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	13/12/2022	13/12/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	13/12/2022	13/12/2022
					CT145	Cardiovascular devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	13/12/2022	13/12/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	13/12/2022	13/12/2022
					CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	13/12/2022	13/12/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	13/12/2022	13/12/2022
					CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	13/12/2022	13/12/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	13/12/2022	13/12/2022
					CT233	Surgical	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	13/12/2022	13/12/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	13/12/2022	13/12/2022
					CT2406	Implantable	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	13/12/2022	13/12/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	13/12/2022	13/12/2022
					CT2407	Active implantable	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	13/12/2022	13/12/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	13/12/2022	13/12/2022
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	13/12/2022	13/12/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	13/12/2022	13/12/2022
					CT2727	Active implant devices/ancillaries	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	13/12/2022	13/12/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	13/12/2022	13/12/2022
					CT2750	Implantable cardiac electrical stimulators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	13/12/2022	13/12/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	13/12/2022	13/12/2022
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	13/12/2022	13/12/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	13/12/2022	13/12/2022
					CT2985	Active	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	13/12/2022	13/12/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	13/12/2022	13/12/2022
					CT3001	Implantable metal	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	13/12/2022	13/12/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	13/12/2022	13/12/2022
					CT3002	Implantable synthetic polymer	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	13/12/2022	13/12/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	13/12/2022	13/12/2022
					CT3006	Implantable inorganic material	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	13/12/2022	13/12/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	13/12/2022	13/12/2022
					CT314	Electrically powered	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	13/12/2022	13/12/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	13/12/2022	13/12/2022
					CT321	Long-term surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	13/12/2022	13/12/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	13/12/2022	13/12/2022
					CT449	Generators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	13/12/2022	13/12/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	13/12/2022	13/12/2022
					CT981	Single use	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	13/12/2022	13/12/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	13/12/2022	13/12/2022
					CT983	Surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	13/12/2022	13/12/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	13/12/2022	13/12/2022
12145	2023		47265	Dual-chamber implantable pacemaker, rate-	CT141	Electrodes/Leads and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	11/12/2022	11/12/2022

				responsive			F1901	Additional Surgery	A24	No Apparent Device Problem	11/12/2022	11/12/2022
					CT145	Cardiovascular devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	11/12/2022	11/12/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	11/12/2022	11/12/2022
					CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	11/12/2022	11/12/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	11/12/2022	11/12/2022
					CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	11/12/2022	11/12/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	11/12/2022	11/12/2022
					CT233	Surgical	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	11/12/2022	11/12/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	11/12/2022	11/12/2022
					CT2406	Implantable	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	11/12/2022	11/12/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	11/12/2022	11/12/2022
					CT2407	Active implantable	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	11/12/2022	11/12/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	11/12/2022	11/12/2022
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	11/12/2022	11/12/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	11/12/2022	11/12/2022
					CT2727	Active implant devices/ancillaries	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	11/12/2022	11/12/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	11/12/2022	11/12/2022
					CT2750	Implantable cardiac electrical stimulators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	11/12/2022	11/12/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	11/12/2022	11/12/2022
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	11/12/2022	11/12/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	11/12/2022	11/12/2022
					CT2985	Active	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	11/12/2022	11/12/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	11/12/2022	11/12/2022
					CT3001	Implantable metal	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	11/12/2022	11/12/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	11/12/2022	11/12/2022
					CT3002	Implantable synthetic polymer	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	11/12/2022	11/12/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	11/12/2022	11/12/2022
					CT3006	Implantable inorganic material	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	11/12/2022	11/12/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	11/12/2022	11/12/2022
					CT314	Electrically powered	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	11/12/2022	11/12/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	11/12/2022	11/12/2022
					CT321	Long-term surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	11/12/2022	11/12/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	11/12/2022	11/12/2022
					CT449	Generators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	11/12/2022	11/12/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	11/12/2022	11/12/2022
					CT981	Single use	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	11/12/2022	11/12/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	11/12/2022	11/12/2022
					CT983	Surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	11/12/2022	11/12/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	11/12/2022	11/12/2022
12146	2023	All other reportable incidents	47265	Dual-chamber implantable pacemaker, rate-responsive	CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem	25/04/2023	25/04/2023
									A070504	Premature Discharge of Battery	25/04/2023	25/04/2023
							F08	Hospitalization or Prolonged Hospitalization	A0705	Battery Problem	25/04/2023	25/04/2023
									A070504	Premature Discharge of Battery	25/04/2023	25/04/2023
							F1905	Device Revision or Replacement	A0705	Battery Problem	25/04/2023	25/04/2023
									A070504	Premature Discharge of Battery	25/04/2023	25/04/2023
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem	25/04/2023	25/04/2023
									A070504	Premature Discharge of Battery	25/04/2023	25/04/2023
							F08	Hospitalization or Prolonged Hospitalization	A0705	Battery Problem	25/04/2023	25/04/2023
									A070504	Premature Discharge of Battery	25/04/2023	25/04/2023
							F1905	Device Revision or Replacement	A0705	Battery Problem	25/04/2023	25/04/2023
									A070504	Premature Discharge of Battery	25/04/2023	25/04/2023
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem	25/04/2023	25/04/2023
									A070504	Premature Discharge of Battery	25/04/2023	25/04/2023
							F08	Hospitalization or Prolonged Hospitalization	A0705	Battery Problem	25/04/2023	25/04/2023
									A070504	Premature Discharge of Battery	25/04/2023	25/04/2023
							F1905	Device Revision or Replacement	A0705	Battery Problem	25/04/2023	25/04/2023
									A070504	Premature Discharge of Battery	25/04/2023	25/04/2023
12147	2023		47265	Dual-chamber implantable pacemaker, rate-responsive	CT141	Electrodes/Leads and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	20/01/2023	20/01/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	20/01/2023	20/01/2023
					CT145	Cardiovascular devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	20/01/2023	20/01/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	20/01/2023	20/01/2023
					CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	20/01/2023	20/01/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	20/01/2023	20/01/2023
					CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	20/01/2023	20/01/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	20/01/2023	20/01/2023
					CT233	Surgical	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	20/01/2023	20/01/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	20/01/2023	20/01/2023
					CT2406	Implantable	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	20/01/2023	20/01/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	20/01/2023	20/01/2023
					CT2407	Active implantable	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	20/01/2023	20/01/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	20/01/2023	20/01/2023
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	20/01/2023	20/01/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	20/01/2023	20/01/2023
					CT2727	Active implant devices/ancillaries	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	20/01/2023	20/01/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	20/01/2023	20/01/2023
					CT2750	Implantable cardiac electrical stimulators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	20/01/2023	20/01/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	20/01/2023	20/01/2023
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	20/01/2023	20/01/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	20/01/2023	20/01/2023
					CT2985	Active	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	20/01/2023	20/01/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	20/01/2023	20/01/2023
					CT3001	Implantable metal	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	20/01/2023	20/01/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	20/01/2023	20/01/2023
					CT3002	Implantable synthetic polymer	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	20/01/2023	20/01/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	20/01/2023	20/01/2023
					CT3006	Implantable inorganic material	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	20/01/2023	20/01/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	20/01/2023	20/01/2023
					CT314	Electrically powered	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	20/01/2023	20/01/2023

							F1901	Additional Surgery	A24	No Apparent Device Problem	20/01/2023	20/01/2023
					CT321	Long-term surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	20/01/2023	20/01/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	20/01/2023	20/01/2023
					CT449	Generators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	20/01/2023	20/01/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	20/01/2023	20/01/2023
					CT981	Single use	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	20/01/2023	20/01/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	20/01/2023	20/01/2023
					CT983	Surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	20/01/2023	20/01/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	20/01/2023	20/01/2023
12148	2023		47265	Dual-chamber implantable pacemaker, rate-responsive	CT141	Electrodes/Leads and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	23/01/2023	23/01/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	23/01/2023	23/01/2023
					CT145	Cardiovascular devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	23/01/2023	23/01/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	23/01/2023	23/01/2023
					CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	23/01/2023	23/01/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	23/01/2023	23/01/2023
					CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	23/01/2023	23/01/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	23/01/2023	23/01/2023
					CT233	Surgical	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	23/01/2023	23/01/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	23/01/2023	23/01/2023
					CT2406	Implantable	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	23/01/2023	23/01/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	23/01/2023	23/01/2023
					CT2407	Active implantable	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	23/01/2023	23/01/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	23/01/2023	23/01/2023
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	23/01/2023	23/01/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	23/01/2023	23/01/2023
					CT2727	Active implant devices/ancillaries	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	23/01/2023	23/01/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	23/01/2023	23/01/2023
					CT2750	Implantable cardiac electrical stimulators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	23/01/2023	23/01/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	23/01/2023	23/01/2023
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	23/01/2023	23/01/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	23/01/2023	23/01/2023
					CT2985	Active	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	23/01/2023	23/01/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	23/01/2023	23/01/2023
					CT3001	Implantable metal	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	23/01/2023	23/01/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	23/01/2023	23/01/2023
					CT3002	Implantable synthetic polymer	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	23/01/2023	23/01/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	23/01/2023	23/01/2023
					CT3006	Implantable inorganic material	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	23/01/2023	23/01/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	23/01/2023	23/01/2023
					CT314	Electrically powered	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	23/01/2023	23/01/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	23/01/2023	23/01/2023
					CT321	Long-term surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	23/01/2023	23/01/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	23/01/2023	23/01/2023
					CT449	Generators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	23/01/2023	23/01/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	23/01/2023	23/01/2023
					CT981	Single use	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	23/01/2023	23/01/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	23/01/2023	23/01/2023
					CT983	Surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	23/01/2023	23/01/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	23/01/2023	23/01/2023
12149	2023	All other reportable incidents	37265	Dual-chamber implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E0112	Dizziness	A071202	Inaccurate Synchronization		
							E0623	Unspecified Heart Problem	A071202	Inaccurate Synchronization		
							F26	No Health Consequences or Impact	A071202	Inaccurate Synchronization		
12150	2023	All other reportable incidents	35852	Single-chamber implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
									A0908	Incorrect, Inadequate or Imprecise Result or Readings		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
									A0908	Incorrect, Inadequate or Imprecise Result or Readings		
12151	2023		47265	Dual-chamber implantable pacemaker, rate-responsive	CT141	Electrodes/Leads and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	23/01/2023	23/01/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	23/01/2023	23/01/2023
					CT145	Cardiovascular devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	23/01/2023	23/01/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	23/01/2023	23/01/2023
					CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	23/01/2023	23/01/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	23/01/2023	23/01/2023
					CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	23/01/2023	23/01/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	23/01/2023	23/01/2023
					CT233	Surgical	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	23/01/2023	23/01/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	23/01/2023	23/01/2023
					CT2406	Implantable	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	23/01/2023	23/01/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	23/01/2023	23/01/2023
					CT2407	Active implantable	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	23/01/2023	23/01/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	23/01/2023	23/01/2023
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	23/01/2023	23/01/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	23/01/2023	23/01/2023
					CT2727	Active implant devices/ancillaries	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	23/01/2023	23/01/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	23/01/2023	23/01/2023
					CT2750	Implantable cardiac electrical stimulators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	23/01/2023	23/01/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	23/01/2023	23/01/2023
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	23/01/2023	23/01/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	23/01/2023	23/01/2023
					CT2985	Active	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	23/01/2023	23/01/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	23/01/2023	23/01/2023
					CT3001	Implantable metal	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	23/01/2023	23/01/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	23/01/2023	23/01/2023
					CT3002	Implantable synthetic polymer	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	23/01/2023	23/01/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	23/01/2023	23/01/2023
					CT3006	Implantable inorganic material	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	23/01/2023	23/01/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	23/01/2023	23/01/2023

					CT314	Electrically powered	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	23/01/2023	23/01/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	23/01/2023	23/01/2023
					CT321	Long-term surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	23/01/2023	23/01/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	23/01/2023	23/01/2023
					CT449	Generators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	23/01/2023	23/01/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	23/01/2023	23/01/2023
					CT981	Single use	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	23/01/2023	23/01/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	23/01/2023	23/01/2023
					CT983	Surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	23/01/2023	23/01/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	23/01/2023	23/01/2023
12152	2023	All other reportable incidents	36241	Pacemaker/defibrillator lead	CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A040101	Fracture	19/01/2023	19/01/2023
									A070101	Failure to Capture	19/01/2023	19/01/2023
									A070102	High Capture Threshold	19/01/2023	19/01/2023
									A072201	High Impedance	19/01/2023	19/01/2023
									A090801	Signal Artifact/Noise	19/01/2023	19/01/2023
							F08	Hospitalization or Prolonged Hospitalization	A040101	Fracture	19/01/2023	19/01/2023
									A070101	Failure to Capture	19/01/2023	19/01/2023
									A070102	High Capture Threshold	19/01/2023	19/01/2023
									A072201	High Impedance	19/01/2023	19/01/2023
									A090801	Signal Artifact/Noise	19/01/2023	19/01/2023
							F1908	Prolonged Surgery	A040101	Fracture	19/01/2023	19/01/2023
									A070101	Failure to Capture	19/01/2023	19/01/2023
									A070102	High Capture Threshold	19/01/2023	19/01/2023
									A072201	High Impedance	19/01/2023	19/01/2023
									A090801	Signal Artifact/Noise	19/01/2023	19/01/2023
12153	2023	All other reportable incidents	35223	Endocardial/interventricular septal pacing lead	CT1684	Cardiac pacemakers and associated devices	E0601	Arrhythmia	A070104	Unstable Capture Threshold		
									A070909	Oversensing		
							F11	Non-Serious Injury/ Illness/ Impairment	A070104	Unstable Capture Threshold		
									A070909	Oversensing		
12154	2023		47265	Dual-chamber implantable pacemaker, rate-responsive	CT141	Electrodes/Leads and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	06/12/2022	06/12/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	06/12/2022	06/12/2022
					CT145	Cardiovascular devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	06/12/2022	06/12/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	06/12/2022	06/12/2022
					CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	06/12/2022	06/12/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	06/12/2022	06/12/2022
					CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	06/12/2022	06/12/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	06/12/2022	06/12/2022
					CT233	Surgical	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	06/12/2022	06/12/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	06/12/2022	06/12/2022
					CT2406	Implantable	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	06/12/2022	06/12/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	06/12/2022	06/12/2022
					CT2407	Active implantable	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	06/12/2022	06/12/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	06/12/2022	06/12/2022
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	06/12/2022	06/12/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	06/12/2022	06/12/2022
					CT2727	Active implant devices/ancillaries	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	06/12/2022	06/12/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	06/12/2022	06/12/2022
					CT2750	Implantable cardiac electrical stimulators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	06/12/2022	06/12/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	06/12/2022	06/12/2022
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	06/12/2022	06/12/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	06/12/2022	06/12/2022
					CT2985	Active	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	06/12/2022	06/12/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	06/12/2022	06/12/2022
					CT3001	Implantable metal	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	06/12/2022	06/12/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	06/12/2022	06/12/2022
					CT3002	Implantable synthetic polymer	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	06/12/2022	06/12/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	06/12/2022	06/12/2022
					CT3006	Implantable inorganic material	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	06/12/2022	06/12/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	06/12/2022	06/12/2022
					CT314	Electrically powered	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	06/12/2022	06/12/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	06/12/2022	06/12/2022
					CT321	Long-term surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	06/12/2022	06/12/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	06/12/2022	06/12/2022
					CT449	Generators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	06/12/2022	06/12/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	06/12/2022	06/12/2022
					CT981	Single use	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	06/12/2022	06/12/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	06/12/2022	06/12/2022
					CT983	Surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	06/12/2022	06/12/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	06/12/2022	06/12/2022
12155	2023	All other reportable incidents	35223	Endocardial/interventricular septal pacing lead	CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070909	Oversensing	08/03/2022	08/03/2022
							F1901	Additional Surgery	A070909	Oversensing	08/03/2022	08/03/2022
12156	2023	All other reportable incidents	47265	Dual-chamber implantable pacemaker, rate-responsive	CT141	Electrodes/Leads and associated devices	E2311	Discomfort	A071102	Failure to Interrogate	25/01/2023	25/01/2023
									A071205	Pacemaker Found in Back-Up Mode	25/01/2023	25/01/2023
									A090407	Therapy Delivered to Incorrect Body Area	25/01/2023	25/01/2023
							F1901	Additional Surgery	A071102	Failure to Interrogate	25/01/2023	25/01/2023
									A071205	Pacemaker Found in Back-Up Mode	25/01/2023	25/01/2023
									A090407	Therapy Delivered to Incorrect Body Area	25/01/2023	25/01/2023
					CT145	Cardiovascular devices	E2311	Discomfort	A071102	Failure to Interrogate	25/01/2023	25/01/2023
									A071205	Pacemaker Found in Back-Up Mode	25/01/2023	25/01/2023
									A090407	Therapy Delivered to Incorrect Body Area	25/01/2023	25/01/2023
							F1901	Additional Surgery	A071102	Failure to Interrogate	25/01/2023	25/01/2023
									A071205	Pacemaker Found in Back-Up Mode	25/01/2023	25/01/2023
									A090407	Therapy Delivered to Incorrect Body Area	25/01/2023	25/01/2023
					CT1684	Cardiac pacemakers and associated devices	E2311	Discomfort	A071102	Failure to Interrogate	25/01/2023	25/01/2023
									A071205	Pacemaker Found in Back-Up Mode	25/01/2023	25/01/2023
									A090407	Therapy Delivered to Incorrect Body Area	25/01/2023	25/01/2023
							F1901	Additional Surgery	A071102	Failure to Interrogate	25/01/2023	25/01/2023
									A071205	Pacemaker Found in Back-Up Mode	25/01/2023	25/01/2023

								A090407	Therapy Delivered to Incorrect Body Area	25/01/2023	25/01/2023
				CT1685	Cardiac pacemakers	E2311	Discomfort	A071102	Failure to Interrogate	25/01/2023	25/01/2023
								A071205	Pacemaker Found in Back-Up Mode	25/01/2023	25/01/2023
								A090407	Therapy Delivered to Incorrect Body Area	25/01/2023	25/01/2023
						F1901	Additional Surgery	A071102	Failure to Interrogate	25/01/2023	25/01/2023
								A071205	Pacemaker Found in Back-Up Mode	25/01/2023	25/01/2023
								A090407	Therapy Delivered to Incorrect Body Area	25/01/2023	25/01/2023
				CT233	Surgical	E2311	Discomfort	A071102	Failure to Interrogate	25/01/2023	25/01/2023
								A071205	Pacemaker Found in Back-Up Mode	25/01/2023	25/01/2023
								A090407	Therapy Delivered to Incorrect Body Area	25/01/2023	25/01/2023
						F1901	Additional Surgery	A071102	Failure to Interrogate	25/01/2023	25/01/2023
								A071205	Pacemaker Found in Back-Up Mode	25/01/2023	25/01/2023
								A090407	Therapy Delivered to Incorrect Body Area	25/01/2023	25/01/2023
				CT2406	Implantable	E2311	Discomfort	A071102	Failure to Interrogate	25/01/2023	25/01/2023
								A071205	Pacemaker Found in Back-Up Mode	25/01/2023	25/01/2023
								A090407	Therapy Delivered to Incorrect Body Area	25/01/2023	25/01/2023
						F1901	Additional Surgery	A071102	Failure to Interrogate	25/01/2023	25/01/2023
								A071205	Pacemaker Found in Back-Up Mode	25/01/2023	25/01/2023
								A090407	Therapy Delivered to Incorrect Body Area	25/01/2023	25/01/2023
				CT2407	Active implantable	E2311	Discomfort	A071102	Failure to Interrogate	25/01/2023	25/01/2023
								A071205	Pacemaker Found in Back-Up Mode	25/01/2023	25/01/2023
								A090407	Therapy Delivered to Incorrect Body Area	25/01/2023	25/01/2023
						F1901	Additional Surgery	A071102	Failure to Interrogate	25/01/2023	25/01/2023
								A071205	Pacemaker Found in Back-Up Mode	25/01/2023	25/01/2023
								A090407	Therapy Delivered to Incorrect Body Area	25/01/2023	25/01/2023
				CT257	Implantable cardiac pacemakers	E2311	Discomfort	A071102	Failure to Interrogate	25/01/2023	25/01/2023
								A071205	Pacemaker Found in Back-Up Mode	25/01/2023	25/01/2023
								A090407	Therapy Delivered to Incorrect Body Area	25/01/2023	25/01/2023
						F1901	Additional Surgery	A071102	Failure to Interrogate	25/01/2023	25/01/2023
								A071205	Pacemaker Found in Back-Up Mode	25/01/2023	25/01/2023
								A090407	Therapy Delivered to Incorrect Body Area	25/01/2023	25/01/2023
				CT2727	Active implant devices/ancillaries	E2311	Discomfort	A071102	Failure to Interrogate	25/01/2023	25/01/2023
								A071205	Pacemaker Found in Back-Up Mode	25/01/2023	25/01/2023
								A090407	Therapy Delivered to Incorrect Body Area	25/01/2023	25/01/2023
						F1901	Additional Surgery	A071102	Failure to Interrogate	25/01/2023	25/01/2023
								A071205	Pacemaker Found in Back-Up Mode	25/01/2023	25/01/2023
								A090407	Therapy Delivered to Incorrect Body Area	25/01/2023	25/01/2023
				CT2750	Implantable cardiac electrical stimulators and associated devices	E2311	Discomfort	A071102	Failure to Interrogate	25/01/2023	25/01/2023
								A071205	Pacemaker Found in Back-Up Mode	25/01/2023	25/01/2023
								A090407	Therapy Delivered to Incorrect Body Area	25/01/2023	25/01/2023
						F1901	Additional Surgery	A071102	Failure to Interrogate	25/01/2023	25/01/2023
								A071205	Pacemaker Found in Back-Up Mode	25/01/2023	25/01/2023
								A090407	Therapy Delivered to Incorrect Body Area	25/01/2023	25/01/2023
				CT2866	Implantable cardiac electrical stimulators	E2311	Discomfort	A071102	Failure to Interrogate	25/01/2023	25/01/2023
								A071205	Pacemaker Found in Back-Up Mode	25/01/2023	25/01/2023
								A090407	Therapy Delivered to Incorrect Body Area	25/01/2023	25/01/2023
						F1901	Additional Surgery	A071102	Failure to Interrogate	25/01/2023	25/01/2023
								A071205	Pacemaker Found in Back-Up Mode	25/01/2023	25/01/2023
								A090407	Therapy Delivered to Incorrect Body Area	25/01/2023	25/01/2023
				CT2985	Active	E2311	Discomfort	A071102	Failure to Interrogate	25/01/2023	25/01/2023
								A071205	Pacemaker Found in Back-Up Mode	25/01/2023	25/01/2023
								A090407	Therapy Delivered to Incorrect Body Area	25/01/2023	25/01/2023
						F1901	Additional Surgery	A071102	Failure to Interrogate	25/01/2023	25/01/2023
								A071205	Pacemaker Found in Back-Up Mode	25/01/2023	25/01/2023
								A090407	Therapy Delivered to Incorrect Body Area	25/01/2023	25/01/2023
				CT3001	Implantable metal	E2311	Discomfort	A071102	Failure to Interrogate	25/01/2023	25/01/2023
								A071205	Pacemaker Found in Back-Up Mode	25/01/2023	25/01/2023
								A090407	Therapy Delivered to Incorrect Body Area	25/01/2023	25/01/2023
						F1901	Additional Surgery	A071102	Failure to Interrogate	25/01/2023	25/01/2023
								A071205	Pacemaker Found in Back-Up Mode	25/01/2023	25/01/2023
								A090407	Therapy Delivered to Incorrect Body Area	25/01/2023	25/01/2023
				CT3002	Implantable synthetic polymer	E2311	Discomfort	A071102	Failure to Interrogate	25/01/2023	25/01/2023
								A071205	Pacemaker Found in Back-Up Mode	25/01/2023	25/01/2023
								A090407	Therapy Delivered to Incorrect Body Area	25/01/2023	25/01/2023
						F1901	Additional Surgery	A071102	Failure to Interrogate	25/01/2023	25/01/2023
								A071205	Pacemaker Found in Back-Up Mode	25/01/2023	25/01/2023
								A090407	Therapy Delivered to Incorrect Body Area	25/01/2023	25/01/2023
				CT3006	Implantable inorganic material	E2311	Discomfort	A071102	Failure to Interrogate	25/01/2023	25/01/2023
								A071205	Pacemaker Found in Back-Up Mode	25/01/2023	25/01/2023
								A090407	Therapy Delivered to Incorrect Body Area	25/01/2023	25/01/2023
						F1901	Additional Surgery	A071102	Failure to Interrogate	25/01/2023	25/01/2023
								A071205	Pacemaker Found in Back-Up Mode	25/01/2023	25/01/2023
								A090407	Therapy Delivered to Incorrect Body Area	25/01/2023	25/01/2023
				CT314	Electrically powered	E2311	Discomfort	A071102	Failure to Interrogate	25/01/2023	25/01/2023
								A071205	Pacemaker Found in Back-Up Mode	25/01/2023	25/01/2023
								A090407	Therapy Delivered to Incorrect Body Area	25/01/2023	25/01/2023
						F1901	Additional Surgery	A071102	Failure to Interrogate	25/01/2023	25/01/2023
								A071205	Pacemaker Found in Back-Up Mode	25/01/2023	25/01/2023
								A090407	Therapy Delivered to Incorrect Body Area	25/01/2023	25/01/2023
				CT321	Long-term surgical invasive	E2311	Discomfort	A071102	Failure to Interrogate	25/01/2023	25/01/2023
								A071205	Pacemaker Found in Back-Up Mode	25/01/2023	25/01/2023
								A090407	Therapy Delivered to Incorrect Body Area	25/01/2023	25/01/2023
						F1901	Additional Surgery	A071102	Failure to Interrogate	25/01/2023	25/01/2023
								A071205	Pacemaker Found in Back-Up Mode	25/01/2023	25/01/2023
								A090407	Therapy Delivered to Incorrect Body Area	25/01/2023	25/01/2023
				CT449	Generators and associated devices	E2311	Discomfort	A071102	Failure to Interrogate	25/01/2023	25/01/2023
								A071205	Pacemaker Found in Back-Up Mode	25/01/2023	25/01/2023
								A090407	Therapy Delivered to Incorrect Body Area	25/01/2023	25/01/2023

							F1901	Additional Surgery	A071102	Failure to Interrogate	25/01/2023	25/01/2023
									A071205	Pacemaker Found in Back-Up Mode	25/01/2023	25/01/2023
									A090407	Therapy Delivered to Incorrect Body Area	25/01/2023	25/01/2023
					CT981	Single use	E2311	Discomfort	A071102	Failure to Interrogate	25/01/2023	25/01/2023
									A071205	Pacemaker Found in Back-Up Mode	25/01/2023	25/01/2023
									A090407	Therapy Delivered to Incorrect Body Area	25/01/2023	25/01/2023
							F1901	Additional Surgery	A071102	Failure to Interrogate	25/01/2023	25/01/2023
									A071205	Pacemaker Found in Back-Up Mode	25/01/2023	25/01/2023
									A090407	Therapy Delivered to Incorrect Body Area	25/01/2023	25/01/2023
					CT983	Surgical invasive	E2311	Discomfort	A071102	Failure to Interrogate	25/01/2023	25/01/2023
									A071205	Pacemaker Found in Back-Up Mode	25/01/2023	25/01/2023
									A090407	Therapy Delivered to Incorrect Body Area	25/01/2023	25/01/2023
							F1901	Additional Surgery	A071102	Failure to Interrogate	25/01/2023	25/01/2023
									A071205	Pacemaker Found in Back-Up Mode	25/01/2023	25/01/2023
									A090407	Therapy Delivered to Incorrect Body Area	25/01/2023	25/01/2023
12157	2023	All other reportable incidents	47265	Dual-chamber implantable pacemaker, rate-responsive	CT141	Electrodes/Leads and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem		
							F1901	Additional Surgery	A24	No Apparent Device Problem		
					CT145	Cardiovascular devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem		
							F1901	Additional Surgery	A24	No Apparent Device Problem		
					CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem		
							F1901	Additional Surgery	A24	No Apparent Device Problem		
					CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem		
							F1901	Additional Surgery	A24	No Apparent Device Problem		
					CT233	Surgical	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem		
							F1901	Additional Surgery	A24	No Apparent Device Problem		
					CT2406	Implantable	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem		
							F1901	Additional Surgery	A24	No Apparent Device Problem		
					CT2407	Active implantable	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem		
							F1901	Additional Surgery	A24	No Apparent Device Problem		
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem		
							F1901	Additional Surgery	A24	No Apparent Device Problem		
					CT2727	Active implant devices/ancillaries	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem		
							F1901	Additional Surgery	A24	No Apparent Device Problem		
					CT2750	Implantable cardiac electrical stimulators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem		
							F1901	Additional Surgery	A24	No Apparent Device Problem		
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem		
							F1901	Additional Surgery	A24	No Apparent Device Problem		
					CT2985	Active	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem		
							F1901	Additional Surgery	A24	No Apparent Device Problem		
					CT3001	Implantable metal	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem		
							F1901	Additional Surgery	A24	No Apparent Device Problem		
					CT3002	Implantable synthetic polymer	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem		
							F1901	Additional Surgery	A24	No Apparent Device Problem		
					CT3006	Implantable inorganic material	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem		
							F1901	Additional Surgery	A24	No Apparent Device Problem		
					CT314	Electrically powered	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem		
							F1901	Additional Surgery	A24	No Apparent Device Problem		
					CT321	Long-term surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem		
							F1901	Additional Surgery	A24	No Apparent Device Problem		
					CT449	Generators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem		
							F1901	Additional Surgery	A24	No Apparent Device Problem		
					CT981	Single use	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem		
							F1901	Additional Surgery	A24	No Apparent Device Problem		
					CT983	Surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem		
							F1901	Additional Surgery	A24	No Apparent Device Problem		
12158	2023	All other reportable incidents	37265	Dual-chamber implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A072201	High Impedance		
							F26	No Health Consequences or Impact	A072201	High Impedance		
12159	2023	All other reportable incidents	35852	Single-chamber implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem	23/01/2023	23/01/2023
									A070504	Premature Discharge of Battery	23/01/2023	23/01/2023
							F08	Hospitalization or Prolonged Hospitalization	A0705	Battery Problem	23/01/2023	23/01/2023
									A070504	Premature Discharge of Battery	23/01/2023	23/01/2023
							F1905	Device Revision or Replacement	A0705	Battery Problem	23/01/2023	23/01/2023
									A070504	Premature Discharge of Battery	23/01/2023	23/01/2023
12160	2023		47265	Dual-chamber implantable pacemaker, rate-responsive	CT141	Electrodes/Leads and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	16/12/2022	16/12/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	16/12/2022	16/12/2022
					CT145	Cardiovascular devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	16/12/2022	16/12/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	16/12/2022	16/12/2022
					CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	16/12/2022	16/12/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	16/12/2022	16/12/2022
					CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	16/12/2022	16/12/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	16/12/2022	16/12/2022
					CT233	Surgical	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	16/12/2022	16/12/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	16/12/2022	16/12/2022
					CT2406	Implantable	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	16/12/2022	16/12/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	16/12/2022	16/12/2022
					CT2407	Active implantable	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	16/12/2022	16/12/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	16/12/2022	16/12/2022
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	16/12/2022	16/12/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	16/12/2022	16/12/2022
					CT2727	Active implant devices/ancillaries	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	16/12/2022	16/12/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	16/12/2022	16/12/2022
					CT2750	Implantable cardiac electrical stimulators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	16/12/2022	16/12/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	16/12/2022	16/12/2022
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	16/12/2022	16/12/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	16/12/2022	16/12/2022
					CT2985	Active	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	16/12/2022	16/12/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	16/12/2022	16/12/2022
					CT3001	Implantable metal	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	16/12/2022	16/12/2022

						F1901	Additional Surgery	A24	No Apparent Device Problem	16/12/2022	16/12/2022
					CT3002	Implantable synthetic polymer	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	16/12/2022	16/12/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	16/12/2022	16/12/2022
					CT3006	Implantable inorganic material	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	16/12/2022	16/12/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	16/12/2022	16/12/2022
					CT314	Electrically powered	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	16/12/2022	16/12/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	16/12/2022	16/12/2022
					CT321	Long-term surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	16/12/2022	16/12/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	16/12/2022	16/12/2022
					CT449	Generators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	16/12/2022	16/12/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	16/12/2022	16/12/2022
					CT981	Single use	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	16/12/2022	16/12/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	16/12/2022	16/12/2022
					CT983	Surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	16/12/2022	16/12/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	16/12/2022	16/12/2022
12161	2023	All other reportable incidents	47265	Dual-chamber implantable pacemaker, rate-responsive	CT141	Electrodes/Leads and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	29/11/2022	29/11/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	29/11/2022	29/11/2022
					CT145	Cardiovascular devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	29/11/2022	29/11/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	29/11/2022	29/11/2022
					CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	29/11/2022	29/11/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	29/11/2022	29/11/2022
					CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	29/11/2022	29/11/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	29/11/2022	29/11/2022
					CT233	Surgical	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	29/11/2022	29/11/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	29/11/2022	29/11/2022
					CT2406	Implantable	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	29/11/2022	29/11/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	29/11/2022	29/11/2022
					CT2407	Active implantable	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	29/11/2022	29/11/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	29/11/2022	29/11/2022
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	29/11/2022	29/11/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	29/11/2022	29/11/2022
					CT2727	Active implant devices/ancillaries	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	29/11/2022	29/11/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	29/11/2022	29/11/2022
					CT2750	Implantable cardiac electrical stimulators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	29/11/2022	29/11/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	29/11/2022	29/11/2022
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	29/11/2022	29/11/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	29/11/2022	29/11/2022
					CT2985	Active	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	29/11/2022	29/11/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	29/11/2022	29/11/2022
					CT3001	Implantable metal	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	29/11/2022	29/11/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	29/11/2022	29/11/2022
					CT3002	Implantable synthetic polymer	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	29/11/2022	29/11/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	29/11/2022	29/11/2022
					CT3006	Implantable inorganic material	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	29/11/2022	29/11/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	29/11/2022	29/11/2022
					CT314	Electrically powered	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	29/11/2022	29/11/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	29/11/2022	29/11/2022
					CT321	Long-term surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	29/11/2022	29/11/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	29/11/2022	29/11/2022
					CT449	Generators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	29/11/2022	29/11/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	29/11/2022	29/11/2022
					CT981	Single use	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	29/11/2022	29/11/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	29/11/2022	29/11/2022
					CT983	Surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	29/11/2022	29/11/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	29/11/2022	29/11/2022
12162	2023	All other reportable incidents	47265	Dual-chamber implantable pacemaker, rate-responsive	CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	15/03/2023	15/03/2023
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	15/03/2023	15/03/2023
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	15/03/2023	15/03/2023
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	15/03/2023	15/03/2023
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	15/03/2023	15/03/2023
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	15/03/2023	15/03/2023
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	15/03/2023	15/03/2023
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	15/03/2023	15/03/2023
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	15/03/2023	15/03/2023
12163	2023	All other reportable incidents	37265	Dual-chamber implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A070101	Failure to Capture	08/11/2022	08/11/2022
									A070908	Failure to Sense	08/11/2022	08/11/2022
									A072201	High Impedance	08/11/2022	08/11/2022
							F1901	Additional Surgery	A070101	Failure to Capture	08/11/2022	08/11/2022
									A070908	Failure to Sense	08/11/2022	08/11/2022
									A072201	High Impedance	08/11/2022	08/11/2022
12164	2023		47265	Dual-chamber implantable pacemaker, rate-responsive	CT141	Electrodes/Leads and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	20/01/2023	20/01/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	20/01/2023	20/01/2023
					CT145	Cardiovascular devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	20/01/2023	20/01/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	20/01/2023	20/01/2023
					CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	20/01/2023	20/01/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	20/01/2023	20/01/2023
					CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	20/01/2023	20/01/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	20/01/2023	20/01/2023
					CT233	Surgical	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	20/01/2023	20/01/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	20/01/2023	20/01/2023
					CT2406	Implantable	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	20/01/2023	20/01/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	20/01/2023	20/01/2023
					CT2407	Active implantable	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	20/01/2023	20/01/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	20/01/2023	20/01/2023
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	20/01/2023	20/01/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	20/01/2023	20/01/2023
					CT2727	Active implant devices/ancillaries	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	20/01/2023	20/01/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	20/01/2023	20/01/2023

					CT2750	Implantable cardiac electrical stimulators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	20/01/2023	20/01/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	20/01/2023	20/01/2023
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	20/01/2023	20/01/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	20/01/2023	20/01/2023
					CT2985	Active	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	20/01/2023	20/01/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	20/01/2023	20/01/2023
					CT3001	Implantable metal	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	20/01/2023	20/01/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	20/01/2023	20/01/2023
					CT3002	Implantable synthetic polymer	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	20/01/2023	20/01/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	20/01/2023	20/01/2023
					CT3006	Implantable inorganic material	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	20/01/2023	20/01/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	20/01/2023	20/01/2023
					CT314	Electrically powered	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	20/01/2023	20/01/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	20/01/2023	20/01/2023
					CT321	Long-term surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	20/01/2023	20/01/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	20/01/2023	20/01/2023
					CT449	Generators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	20/01/2023	20/01/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	20/01/2023	20/01/2023
					CT981	Single use	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	20/01/2023	20/01/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	20/01/2023	20/01/2023
					CT983	Surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	20/01/2023	20/01/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	20/01/2023	20/01/2023
12165	2023	All other reportable incidents	47265	Dual-chamber implantable pacemaker, rate-responsive	CT141	Electrodes/Leads and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	24/01/2023	24/01/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	24/01/2023	24/01/2023
					CT145	Cardiovascular devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	24/01/2023	24/01/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	24/01/2023	24/01/2023
					CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	24/01/2023	24/01/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	24/01/2023	24/01/2023
					CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	24/01/2023	24/01/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	24/01/2023	24/01/2023
					CT233	Surgical	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	24/01/2023	24/01/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	24/01/2023	24/01/2023
					CT2406	Implantable	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	24/01/2023	24/01/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	24/01/2023	24/01/2023
					CT2407	Active implantable	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	24/01/2023	24/01/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	24/01/2023	24/01/2023
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	24/01/2023	24/01/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	24/01/2023	24/01/2023
					CT2727	Active implant devices/ancillaries	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	24/01/2023	24/01/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	24/01/2023	24/01/2023
					CT2750	Implantable cardiac electrical stimulators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	24/01/2023	24/01/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	24/01/2023	24/01/2023
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	24/01/2023	24/01/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	24/01/2023	24/01/2023
					CT2985	Active	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	24/01/2023	24/01/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	24/01/2023	24/01/2023
					CT3001	Implantable metal	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	24/01/2023	24/01/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	24/01/2023	24/01/2023
					CT3002	Implantable synthetic polymer	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	24/01/2023	24/01/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	24/01/2023	24/01/2023
					CT3006	Implantable inorganic material	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	24/01/2023	24/01/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	24/01/2023	24/01/2023
					CT314	Electrically powered	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	24/01/2023	24/01/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	24/01/2023	24/01/2023
					CT321	Long-term surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	24/01/2023	24/01/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	24/01/2023	24/01/2023
					CT449	Generators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	24/01/2023	24/01/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	24/01/2023	24/01/2023
					CT981	Single use	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	24/01/2023	24/01/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	24/01/2023	24/01/2023
					CT983	Surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	24/01/2023	24/01/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	24/01/2023	24/01/2023
12166	2023	All other reportable incidents	47265	Dual-chamber implantable pacemaker, rate-responsive	CT141	Electrodes/Leads and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	18/01/2023	18/01/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	18/01/2023	18/01/2023
					CT145	Cardiovascular devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	18/01/2023	18/01/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	18/01/2023	18/01/2023
					CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	18/01/2023	18/01/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	18/01/2023	18/01/2023
					CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	18/01/2023	18/01/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	18/01/2023	18/01/2023
					CT233	Surgical	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	18/01/2023	18/01/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	18/01/2023	18/01/2023
					CT2406	Implantable	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	18/01/2023	18/01/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	18/01/2023	18/01/2023
					CT2407	Active implantable	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	18/01/2023	18/01/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	18/01/2023	18/01/2023
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	18/01/2023	18/01/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	18/01/2023	18/01/2023
					CT2727	Active implant devices/ancillaries	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	18/01/2023	18/01/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	18/01/2023	18/01/2023
					CT2750	Implantable cardiac electrical stimulators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	18/01/2023	18/01/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	18/01/2023	18/01/2023
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	18/01/2023	18/01/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	18/01/2023	18/01/2023
					CT2985	Active	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	18/01/2023	18/01/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	18/01/2023	18/01/2023
					CT3001	Implantable metal	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	18/01/2023	18/01/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	18/01/2023	18/01/2023

					CT3002	Implantable synthetic polymer	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	18/01/2023	18/01/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	18/01/2023	18/01/2023
					CT3006	Implantable inorganic material	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	18/01/2023	18/01/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	18/01/2023	18/01/2023
					CT314	Electrically powered	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	18/01/2023	18/01/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	18/01/2023	18/01/2023
					CT321	Long-term surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	18/01/2023	18/01/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	18/01/2023	18/01/2023
					CT449	Generators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	18/01/2023	18/01/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	18/01/2023	18/01/2023
					CT981	Single use	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	18/01/2023	18/01/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	18/01/2023	18/01/2023
					CT983	Surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	18/01/2023	18/01/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	18/01/2023	18/01/2023
12167	2023		47265	Dual-chamber implantable pacemaker, rate-responsive	CT141	Electrodes/Leads and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem		
							F1901	Additional Surgery	A24	No Apparent Device Problem		
					CT145	Cardiovascular devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem		
							F1901	Additional Surgery	A24	No Apparent Device Problem		
					CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem		
							F1901	Additional Surgery	A24	No Apparent Device Problem		
					CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem		
							F1901	Additional Surgery	A24	No Apparent Device Problem		
					CT233	Surgical	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem		
							F1901	Additional Surgery	A24	No Apparent Device Problem		
					CT2406	Implantable	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem		
							F1901	Additional Surgery	A24	No Apparent Device Problem		
					CT2407	Active implantable	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem		
							F1901	Additional Surgery	A24	No Apparent Device Problem		
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem		
							F1901	Additional Surgery	A24	No Apparent Device Problem		
					CT2727	Active implant devices/ancillaries	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem		
							F1901	Additional Surgery	A24	No Apparent Device Problem		
					CT2750	Implantable cardiac electrical stimulators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem		
							F1901	Additional Surgery	A24	No Apparent Device Problem		
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem		
							F1901	Additional Surgery	A24	No Apparent Device Problem		
					CT2985	Active	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem		
							F1901	Additional Surgery	A24	No Apparent Device Problem		
					CT3001	Implantable metal	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem		
							F1901	Additional Surgery	A24	No Apparent Device Problem		
					CT3002	Implantable synthetic polymer	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem		
							F1901	Additional Surgery	A24	No Apparent Device Problem		
					CT3006	Implantable inorganic material	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem		
							F1901	Additional Surgery	A24	No Apparent Device Problem		
					CT314	Electrically powered	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem		
							F1901	Additional Surgery	A24	No Apparent Device Problem		
					CT321	Long-term surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem		
							F1901	Additional Surgery	A24	No Apparent Device Problem		
					CT449	Generators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem		
							F1901	Additional Surgery	A24	No Apparent Device Problem		
					CT981	Single use	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem		
							F1901	Additional Surgery	A24	No Apparent Device Problem		
					CT983	Surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem		
							F1901	Additional Surgery	A24	No Apparent Device Problem		
12168	2023	All other reportable incidents	37265	Dual-chamber implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	21/02/2023	21/02/2023
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	21/02/2023	21/02/2023
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	21/02/2023	21/02/2023
12169	2023		47265	Dual-chamber implantable pacemaker, rate-responsive	CT141	Electrodes/Leads and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	08/01/2023	08/01/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	08/01/2023	08/01/2023
					CT145	Cardiovascular devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	08/01/2023	08/01/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	08/01/2023	08/01/2023
					CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	08/01/2023	08/01/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	08/01/2023	08/01/2023
					CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	08/01/2023	08/01/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	08/01/2023	08/01/2023
					CT233	Surgical	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	08/01/2023	08/01/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	08/01/2023	08/01/2023
					CT2406	Implantable	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	08/01/2023	08/01/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	08/01/2023	08/01/2023
					CT2407	Active implantable	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	08/01/2023	08/01/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	08/01/2023	08/01/2023
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	08/01/2023	08/01/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	08/01/2023	08/01/2023
					CT2727	Active implant devices/ancillaries	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	08/01/2023	08/01/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	08/01/2023	08/01/2023
					CT2750	Implantable cardiac electrical stimulators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	08/01/2023	08/01/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	08/01/2023	08/01/2023
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	08/01/2023	08/01/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	08/01/2023	08/01/2023
					CT2985	Active	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	08/01/2023	08/01/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	08/01/2023	08/01/2023
					CT3001	Implantable metal	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	08/01/2023	08/01/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	08/01/2023	08/01/2023
					CT3002	Implantable synthetic polymer	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	08/01/2023	08/01/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	08/01/2023	08/01/2023
					CT3006	Implantable inorganic material	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	08/01/2023	08/01/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	08/01/2023	08/01/2023
					CT314	Electrically powered	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	08/01/2023	08/01/2023

							F1901	Additional Surgery	A24	No Apparent Device Problem	08/01/2023	08/01/2023
					CT321	Long-term surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	08/01/2023	08/01/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	08/01/2023	08/01/2023
					CT449	Generators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	08/01/2023	08/01/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	08/01/2023	08/01/2023
					CT981	Single use	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	08/01/2023	08/01/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	08/01/2023	08/01/2023
					CT983	Surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	08/01/2023	08/01/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	08/01/2023	08/01/2023
12170	2023		47265	Dual-chamber implantable pacemaker, rate-responsive	CT141	Electrodes/Leads and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	08/01/2023	08/01/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	08/01/2023	08/01/2023
					CT145	Cardiovascular devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	08/01/2023	08/01/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	08/01/2023	08/01/2023
					CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	08/01/2023	08/01/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	08/01/2023	08/01/2023
					CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	08/01/2023	08/01/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	08/01/2023	08/01/2023
					CT233	Surgical	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	08/01/2023	08/01/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	08/01/2023	08/01/2023
					CT2406	Implantable	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	08/01/2023	08/01/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	08/01/2023	08/01/2023
					CT2407	Active implantable	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	08/01/2023	08/01/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	08/01/2023	08/01/2023
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	08/01/2023	08/01/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	08/01/2023	08/01/2023
					CT2727	Active implant devices/ancillaries	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	08/01/2023	08/01/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	08/01/2023	08/01/2023
					CT2750	Implantable cardiac electrical stimulators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	08/01/2023	08/01/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	08/01/2023	08/01/2023
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	08/01/2023	08/01/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	08/01/2023	08/01/2023
					CT2985	Active	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	08/01/2023	08/01/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	08/01/2023	08/01/2023
					CT3001	Implantable metal	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	08/01/2023	08/01/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	08/01/2023	08/01/2023
					CT3002	Implantable synthetic polymer	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	08/01/2023	08/01/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	08/01/2023	08/01/2023
					CT3006	Implantable inorganic material	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	08/01/2023	08/01/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	08/01/2023	08/01/2023
					CT314	Electrically powered	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	08/01/2023	08/01/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	08/01/2023	08/01/2023
					CT321	Long-term surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	08/01/2023	08/01/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	08/01/2023	08/01/2023
					CT449	Generators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	08/01/2023	08/01/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	08/01/2023	08/01/2023
					CT981	Single use	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	08/01/2023	08/01/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	08/01/2023	08/01/2023
					CT983	Surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	08/01/2023	08/01/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	08/01/2023	08/01/2023
12171	2023		47265	Dual-chamber implantable pacemaker, rate-responsive	CT141	Electrodes/Leads and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	01/10/2022	31/10/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	01/10/2022	31/10/2022
					CT145	Cardiovascular devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	01/10/2022	31/10/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	01/10/2022	31/10/2022
					CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	01/10/2022	31/10/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	01/10/2022	31/10/2022
					CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	01/10/2022	31/10/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	01/10/2022	31/10/2022
					CT233	Surgical	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	01/10/2022	31/10/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	01/10/2022	31/10/2022
					CT2406	Implantable	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	01/10/2022	31/10/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	01/10/2022	31/10/2022
					CT2407	Active implantable	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	01/10/2022	31/10/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	01/10/2022	31/10/2022
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	01/10/2022	31/10/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	01/10/2022	31/10/2022
					CT2727	Active implant devices/ancillaries	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	01/10/2022	31/10/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	01/10/2022	31/10/2022
					CT2750	Implantable cardiac electrical stimulators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	01/10/2022	31/10/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	01/10/2022	31/10/2022
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	01/10/2022	31/10/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	01/10/2022	31/10/2022
					CT2985	Active	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	01/10/2022	31/10/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	01/10/2022	31/10/2022
					CT3001	Implantable metal	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	01/10/2022	31/10/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	01/10/2022	31/10/2022
					CT3002	Implantable synthetic polymer	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	01/10/2022	31/10/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	01/10/2022	31/10/2022
					CT3006	Implantable inorganic material	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	01/10/2022	31/10/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	01/10/2022	31/10/2022
					CT314	Electrically powered	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	01/10/2022	31/10/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	01/10/2022	31/10/2022
					CT321	Long-term surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	01/10/2022	31/10/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	01/10/2022	31/10/2022
					CT449	Generators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	01/10/2022	31/10/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	01/10/2022	31/10/2022
					CT981	Single use	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	01/10/2022	31/10/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	01/10/2022	31/10/2022
					CT983	Surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	01/10/2022	31/10/2022

							F1901	Additional Surgery	A24	No Apparent Device Problem	01/10/2022	31/10/2022
12172	2023		35223	Endocardial/interventricular septal pacing lead	CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A051201	Device Dislodged or Dislocated	07/12/2022	07/12/2022
									A070909	Oversensing	07/12/2022	07/12/2022
									A072202	Low Impedance	07/12/2022	07/12/2022
							F1901	Additional Surgery	A051201	Device Dislodged or Dislocated	07/12/2022	07/12/2022
									A070909	Oversensing	07/12/2022	07/12/2022
									A072202	Low Impedance	07/12/2022	07/12/2022
12173	2023	All other reportable incidents	47267	Single-chamber implantable pacemaker, rate-responsive	CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A070101	Failure to Capture		
							F26	No Health Consequences or Impact	A070101	Failure to Capture		
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A070101	Failure to Capture		
							F26	No Health Consequences or Impact	A070101	Failure to Capture		
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A070101	Failure to Capture		
							F26	No Health Consequences or Impact	A070101	Failure to Capture		
12174	2023	All other reportable incidents	17882	Physiologic-monitoring defibrillation system	CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A26	Insufficient Device Problem Information		
							F27	Problem Identified During Non-Clinical Procedure	A26	Insufficient Device Problem Information		
12175	2023		47265	Dual-chamber implantable pacemaker, rate-responsive	CT141	Electrodes/Leads and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	08/01/2023	08/01/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	08/01/2023	08/01/2023
					CT145	Cardiovascular devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	08/01/2023	08/01/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	08/01/2023	08/01/2023
					CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	08/01/2023	08/01/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	08/01/2023	08/01/2023
					CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	08/01/2023	08/01/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	08/01/2023	08/01/2023
					CT233	Surgical	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	08/01/2023	08/01/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	08/01/2023	08/01/2023
					CT2406	Implantable	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	08/01/2023	08/01/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	08/01/2023	08/01/2023
					CT2407	Active implantable	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	08/01/2023	08/01/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	08/01/2023	08/01/2023
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	08/01/2023	08/01/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	08/01/2023	08/01/2023
					CT2727	Active implant devices/ancillaries	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	08/01/2023	08/01/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	08/01/2023	08/01/2023
					CT2750	Implantable cardiac electrical stimulators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	08/01/2023	08/01/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	08/01/2023	08/01/2023
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	08/01/2023	08/01/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	08/01/2023	08/01/2023
					CT2985	Active	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	08/01/2023	08/01/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	08/01/2023	08/01/2023
					CT3001	Implantable metal	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	08/01/2023	08/01/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	08/01/2023	08/01/2023
					CT3002	Implantable synthetic polymer	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	08/01/2023	08/01/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	08/01/2023	08/01/2023
					CT3006	Implantable inorganic material	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	08/01/2023	08/01/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	08/01/2023	08/01/2023
					CT314	Electrically powered	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	08/01/2023	08/01/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	08/01/2023	08/01/2023
					CT321	Long-term surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	08/01/2023	08/01/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	08/01/2023	08/01/2023
					CT449	Generators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	08/01/2023	08/01/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	08/01/2023	08/01/2023
					CT981	Single use	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	08/01/2023	08/01/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	08/01/2023	08/01/2023
					CT983	Surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	08/01/2023	08/01/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	08/01/2023	08/01/2023
12176	2023		47265	Dual-chamber implantable pacemaker, rate-responsive	CT141	Electrodes/Leads and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem		
							F1901	Additional Surgery	A24	No Apparent Device Problem		
					CT145	Cardiovascular devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem		
							F1901	Additional Surgery	A24	No Apparent Device Problem		
					CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem		
							F1901	Additional Surgery	A24	No Apparent Device Problem		
					CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem		
							F1901	Additional Surgery	A24	No Apparent Device Problem		
					CT233	Surgical	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem		
							F1901	Additional Surgery	A24	No Apparent Device Problem		
					CT2406	Implantable	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem		
							F1901	Additional Surgery	A24	No Apparent Device Problem		
					CT2407	Active implantable	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem		
							F1901	Additional Surgery	A24	No Apparent Device Problem		
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem		
							F1901	Additional Surgery	A24	No Apparent Device Problem		
					CT2727	Active implant devices/ancillaries	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem		
							F1901	Additional Surgery	A24	No Apparent Device Problem		
					CT2750	Implantable cardiac electrical stimulators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem		
							F1901	Additional Surgery	A24	No Apparent Device Problem		
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem		
							F1901	Additional Surgery	A24	No Apparent Device Problem		
					CT2985	Active	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem		
							F1901	Additional Surgery	A24	No Apparent Device Problem		
					CT3001	Implantable metal	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem		
							F1901	Additional Surgery	A24	No Apparent Device Problem		
					CT3002	Implantable synthetic polymer	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem		
							F1901	Additional Surgery	A24	No Apparent Device Problem		
					CT3006	Implantable inorganic material	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem		
							F1901	Additional Surgery	A24	No Apparent Device Problem		
					CT314	Electrically powered	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem		
							F1901	Additional Surgery	A24	No Apparent Device Problem		
					CT321	Long-term surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem		

						F1901	Additional Surgery	A24	No Apparent Device Problem			
				CT449	Generators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem			
						F1901	Additional Surgery	A24	No Apparent Device Problem			
				CT981	Single use	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem			
						F1901	Additional Surgery	A24	No Apparent Device Problem			
				CT983	Surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem			
						F1901	Additional Surgery	A24	No Apparent Device Problem			
12177	2023		47265	Dual-chamber implantable pacemaker, rate-responsive	CT141	Electrodes/Leads and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	24/01/2023	24/01/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	24/01/2023	24/01/2023
					CT145	Cardiovascular devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	24/01/2023	24/01/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	24/01/2023	24/01/2023
					CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	24/01/2023	24/01/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	24/01/2023	24/01/2023
					CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	24/01/2023	24/01/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	24/01/2023	24/01/2023
					CT233	Surgical	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	24/01/2023	24/01/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	24/01/2023	24/01/2023
					CT2406	Implantable	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	24/01/2023	24/01/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	24/01/2023	24/01/2023
					CT2407	Active implantable	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	24/01/2023	24/01/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	24/01/2023	24/01/2023
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	24/01/2023	24/01/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	24/01/2023	24/01/2023
					CT2727	Active implant devices/ancillaries	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	24/01/2023	24/01/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	24/01/2023	24/01/2023
					CT2750	Implantable cardiac electrical stimulators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	24/01/2023	24/01/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	24/01/2023	24/01/2023
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	24/01/2023	24/01/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	24/01/2023	24/01/2023
					CT2985	Active	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	24/01/2023	24/01/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	24/01/2023	24/01/2023
					CT3001	Implantable metal	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	24/01/2023	24/01/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	24/01/2023	24/01/2023
					CT3002	Implantable synthetic polymer	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	24/01/2023	24/01/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	24/01/2023	24/01/2023
					CT3006	Implantable inorganic material	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	24/01/2023	24/01/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	24/01/2023	24/01/2023
					CT314	Electrically powered	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	24/01/2023	24/01/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	24/01/2023	24/01/2023
					CT321	Long-term surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	24/01/2023	24/01/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	24/01/2023	24/01/2023
					CT449	Generators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	24/01/2023	24/01/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	24/01/2023	24/01/2023
					CT981	Single use	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	24/01/2023	24/01/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	24/01/2023	24/01/2023
					CT983	Surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	24/01/2023	24/01/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	24/01/2023	24/01/2023
12178	2023		47265	Dual-chamber implantable pacemaker, rate-responsive	CT141	Electrodes/Leads and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	07/12/2022	07/12/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	07/12/2022	07/12/2022
					CT145	Cardiovascular devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	07/12/2022	07/12/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	07/12/2022	07/12/2022
					CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	07/12/2022	07/12/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	07/12/2022	07/12/2022
					CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	07/12/2022	07/12/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	07/12/2022	07/12/2022
					CT233	Surgical	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	07/12/2022	07/12/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	07/12/2022	07/12/2022
					CT2406	Implantable	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	07/12/2022	07/12/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	07/12/2022	07/12/2022
					CT2407	Active implantable	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	07/12/2022	07/12/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	07/12/2022	07/12/2022
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	07/12/2022	07/12/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	07/12/2022	07/12/2022
					CT2727	Active implant devices/ancillaries	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	07/12/2022	07/12/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	07/12/2022	07/12/2022
					CT2750	Implantable cardiac electrical stimulators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	07/12/2022	07/12/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	07/12/2022	07/12/2022
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	07/12/2022	07/12/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	07/12/2022	07/12/2022
					CT2985	Active	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	07/12/2022	07/12/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	07/12/2022	07/12/2022
					CT3001	Implantable metal	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	07/12/2022	07/12/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	07/12/2022	07/12/2022
					CT3002	Implantable synthetic polymer	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	07/12/2022	07/12/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	07/12/2022	07/12/2022
					CT3006	Implantable inorganic material	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	07/12/2022	07/12/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	07/12/2022	07/12/2022
					CT314	Electrically powered	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	07/12/2022	07/12/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	07/12/2022	07/12/2022
					CT321	Long-term surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	07/12/2022	07/12/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	07/12/2022	07/12/2022
					CT449	Generators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	07/12/2022	07/12/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	07/12/2022	07/12/2022
					CT981	Single use	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	07/12/2022	07/12/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	07/12/2022	07/12/2022
					CT983	Surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	07/12/2022	07/12/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	07/12/2022	07/12/2022
12179	2023		47265	Dual-chamber implantable pacemaker, rate-	CT141	Electrodes/Leads and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	16/12/2022	16/12/2022

				responsive			F1901	Additional Surgery	A24	No Apparent Device Problem	16/12/2022	16/12/2022
					CT145	Cardiovascular devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	16/12/2022	16/12/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	16/12/2022	16/12/2022
					CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	16/12/2022	16/12/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	16/12/2022	16/12/2022
					CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	16/12/2022	16/12/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	16/12/2022	16/12/2022
					CT233	Surgical	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	16/12/2022	16/12/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	16/12/2022	16/12/2022
					CT2406	Implantable	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	16/12/2022	16/12/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	16/12/2022	16/12/2022
					CT2407	Active implantable	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	16/12/2022	16/12/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	16/12/2022	16/12/2022
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	16/12/2022	16/12/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	16/12/2022	16/12/2022
					CT2727	Active implant devices/ancillaries	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	16/12/2022	16/12/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	16/12/2022	16/12/2022
					CT2750	Implantable cardiac electrical stimulators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	16/12/2022	16/12/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	16/12/2022	16/12/2022
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	16/12/2022	16/12/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	16/12/2022	16/12/2022
					CT2985	Active	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	16/12/2022	16/12/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	16/12/2022	16/12/2022
					CT3001	Implantable metal	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	16/12/2022	16/12/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	16/12/2022	16/12/2022
					CT3002	Implantable synthetic polymer	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	16/12/2022	16/12/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	16/12/2022	16/12/2022
					CT3006	Implantable inorganic material	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	16/12/2022	16/12/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	16/12/2022	16/12/2022
					CT314	Electrically powered	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	16/12/2022	16/12/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	16/12/2022	16/12/2022
					CT321	Long-term surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	16/12/2022	16/12/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	16/12/2022	16/12/2022
					CT449	Generators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	16/12/2022	16/12/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	16/12/2022	16/12/2022
					CT981	Single use	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	16/12/2022	16/12/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	16/12/2022	16/12/2022
					CT983	Surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	16/12/2022	16/12/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	16/12/2022	16/12/2022
12180	2023		47265	Dual-chamber implantable pacemaker, rate-responsive	CT141	Electrodes/Leads and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	24/01/2023	24/01/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	24/01/2023	24/01/2023
					CT145	Cardiovascular devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	24/01/2023	24/01/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	24/01/2023	24/01/2023
					CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	24/01/2023	24/01/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	24/01/2023	24/01/2023
					CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	24/01/2023	24/01/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	24/01/2023	24/01/2023
					CT233	Surgical	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	24/01/2023	24/01/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	24/01/2023	24/01/2023
					CT2406	Implantable	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	24/01/2023	24/01/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	24/01/2023	24/01/2023
					CT2407	Active implantable	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	24/01/2023	24/01/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	24/01/2023	24/01/2023
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	24/01/2023	24/01/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	24/01/2023	24/01/2023
					CT2727	Active implant devices/ancillaries	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	24/01/2023	24/01/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	24/01/2023	24/01/2023
					CT2750	Implantable cardiac electrical stimulators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	24/01/2023	24/01/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	24/01/2023	24/01/2023
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	24/01/2023	24/01/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	24/01/2023	24/01/2023
					CT2985	Active	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	24/01/2023	24/01/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	24/01/2023	24/01/2023
					CT3001	Implantable metal	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	24/01/2023	24/01/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	24/01/2023	24/01/2023
					CT3002	Implantable synthetic polymer	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	24/01/2023	24/01/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	24/01/2023	24/01/2023
					CT3006	Implantable inorganic material	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	24/01/2023	24/01/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	24/01/2023	24/01/2023
					CT314	Electrically powered	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	24/01/2023	24/01/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	24/01/2023	24/01/2023
					CT321	Long-term surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	24/01/2023	24/01/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	24/01/2023	24/01/2023
					CT449	Generators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	24/01/2023	24/01/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	24/01/2023	24/01/2023
					CT981	Single use	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	24/01/2023	24/01/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	24/01/2023	24/01/2023
					CT983	Surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	24/01/2023	24/01/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	24/01/2023	24/01/2023
12181	2023	All other reportable incidents	47265	Dual-chamber implantable pacemaker, rate-responsive	CT141	Electrodes/Leads and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	18/01/2023	18/01/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	18/01/2023	18/01/2023
					CT145	Cardiovascular devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	18/01/2023	18/01/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	18/01/2023	18/01/2023
					CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	18/01/2023	18/01/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	18/01/2023	18/01/2023
					CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	18/01/2023	18/01/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	18/01/2023	18/01/2023
					CT233	Surgical	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	18/01/2023	18/01/2023

							F1901	Additional Surgery	A24	No Apparent Device Problem	18/01/2023	18/01/2023
					CT2406	Implantable	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	18/01/2023	18/01/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	18/01/2023	18/01/2023
					CT2407	Active implantable	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	18/01/2023	18/01/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	18/01/2023	18/01/2023
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	18/01/2023	18/01/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	18/01/2023	18/01/2023
					CT2727	Active implant devices/ancillaries	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	18/01/2023	18/01/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	18/01/2023	18/01/2023
					CT2750	Implantable cardiac electrical stimulators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	18/01/2023	18/01/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	18/01/2023	18/01/2023
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	18/01/2023	18/01/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	18/01/2023	18/01/2023
					CT2985	Active	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	18/01/2023	18/01/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	18/01/2023	18/01/2023
					CT3001	Implantable metal	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	18/01/2023	18/01/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	18/01/2023	18/01/2023
					CT3002	Implantable synthetic polymer	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	18/01/2023	18/01/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	18/01/2023	18/01/2023
					CT3006	Implantable inorganic material	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	18/01/2023	18/01/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	18/01/2023	18/01/2023
					CT314	Electrically powered	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	18/01/2023	18/01/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	18/01/2023	18/01/2023
					CT321	Long-term surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	18/01/2023	18/01/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	18/01/2023	18/01/2023
					CT449	Generators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	18/01/2023	18/01/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	18/01/2023	18/01/2023
					CT981	Single use	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	18/01/2023	18/01/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	18/01/2023	18/01/2023
					CT983	Surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	18/01/2023	18/01/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	18/01/2023	18/01/2023
12182	2023		47265	Dual-chamber implantable pacemaker, rate-responsive	CT141	Electrodes/Leads and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	18/01/2023	18/01/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	18/01/2023	18/01/2023
					CT145	Cardiovascular devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	18/01/2023	18/01/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	18/01/2023	18/01/2023
					CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	18/01/2023	18/01/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	18/01/2023	18/01/2023
					CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	18/01/2023	18/01/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	18/01/2023	18/01/2023
					CT233	Surgical	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	18/01/2023	18/01/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	18/01/2023	18/01/2023
					CT2406	Implantable	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	18/01/2023	18/01/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	18/01/2023	18/01/2023
					CT2407	Active implantable	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	18/01/2023	18/01/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	18/01/2023	18/01/2023
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	18/01/2023	18/01/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	18/01/2023	18/01/2023
					CT2727	Active implant devices/ancillaries	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	18/01/2023	18/01/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	18/01/2023	18/01/2023
					CT2750	Implantable cardiac electrical stimulators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	18/01/2023	18/01/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	18/01/2023	18/01/2023
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	18/01/2023	18/01/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	18/01/2023	18/01/2023
					CT2985	Active	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	18/01/2023	18/01/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	18/01/2023	18/01/2023
					CT3001	Implantable metal	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	18/01/2023	18/01/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	18/01/2023	18/01/2023
					CT3002	Implantable synthetic polymer	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	18/01/2023	18/01/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	18/01/2023	18/01/2023
					CT3006	Implantable inorganic material	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	18/01/2023	18/01/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	18/01/2023	18/01/2023
					CT314	Electrically powered	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	18/01/2023	18/01/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	18/01/2023	18/01/2023
					CT321	Long-term surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	18/01/2023	18/01/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	18/01/2023	18/01/2023
					CT449	Generators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	18/01/2023	18/01/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	18/01/2023	18/01/2023
					CT981	Single use	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	18/01/2023	18/01/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	18/01/2023	18/01/2023
					CT983	Surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	18/01/2023	18/01/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	18/01/2023	18/01/2023
12183	2023		47265	Dual-chamber implantable pacemaker, rate-responsive	CT141	Electrodes/Leads and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	18/01/2023	18/01/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	18/01/2023	18/01/2023
					CT145	Cardiovascular devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	18/01/2023	18/01/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	18/01/2023	18/01/2023
					CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	18/01/2023	18/01/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	18/01/2023	18/01/2023
					CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	18/01/2023	18/01/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	18/01/2023	18/01/2023
					CT233	Surgical	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	18/01/2023	18/01/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	18/01/2023	18/01/2023
					CT2406	Implantable	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	18/01/2023	18/01/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	18/01/2023	18/01/2023
					CT2407	Active implantable	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	18/01/2023	18/01/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	18/01/2023	18/01/2023
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	18/01/2023	18/01/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	18/01/2023	18/01/2023
					CT2727	Active implant devices/ancillaries	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	18/01/2023	18/01/2023

							F1901	Additional Surgery	A24	No Apparent Device Problem	18/01/2023	18/01/2023
					CT2750	Implantable cardiac electrical stimulators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	18/01/2023	18/01/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	18/01/2023	18/01/2023
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	18/01/2023	18/01/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	18/01/2023	18/01/2023
					CT2985	Active	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	18/01/2023	18/01/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	18/01/2023	18/01/2023
					CT3001	Implantable metal	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	18/01/2023	18/01/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	18/01/2023	18/01/2023
					CT3002	Implantable synthetic polymer	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	18/01/2023	18/01/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	18/01/2023	18/01/2023
					CT3006	Implantable inorganic material	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	18/01/2023	18/01/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	18/01/2023	18/01/2023
					CT314	Electrically powered	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	18/01/2023	18/01/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	18/01/2023	18/01/2023
					CT321	Long-term surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	18/01/2023	18/01/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	18/01/2023	18/01/2023
					CT449	Generators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	18/01/2023	18/01/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	18/01/2023	18/01/2023
					CT981	Single use	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	18/01/2023	18/01/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	18/01/2023	18/01/2023
					CT983	Surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	18/01/2023	18/01/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	18/01/2023	18/01/2023
12184	2023		47265	Dual-chamber implantable pacemaker, rate-responsive	CT141	Electrodes/Leads and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	28/11/2022	28/11/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	28/11/2022	28/11/2022
					CT145	Cardiovascular devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	28/11/2022	28/11/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	28/11/2022	28/11/2022
					CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	28/11/2022	28/11/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	28/11/2022	28/11/2022
					CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	28/11/2022	28/11/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	28/11/2022	28/11/2022
					CT233	Surgical	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	28/11/2022	28/11/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	28/11/2022	28/11/2022
					CT2406	Implantable	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	28/11/2022	28/11/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	28/11/2022	28/11/2022
					CT2407	Active implantable	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	28/11/2022	28/11/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	28/11/2022	28/11/2022
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	28/11/2022	28/11/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	28/11/2022	28/11/2022
					CT2727	Active implant devices/ancillaries	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	28/11/2022	28/11/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	28/11/2022	28/11/2022
					CT2750	Implantable cardiac electrical stimulators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	28/11/2022	28/11/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	28/11/2022	28/11/2022
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	28/11/2022	28/11/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	28/11/2022	28/11/2022
					CT2985	Active	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	28/11/2022	28/11/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	28/11/2022	28/11/2022
					CT3001	Implantable metal	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	28/11/2022	28/11/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	28/11/2022	28/11/2022
					CT3002	Implantable synthetic polymer	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	28/11/2022	28/11/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	28/11/2022	28/11/2022
					CT3006	Implantable inorganic material	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	28/11/2022	28/11/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	28/11/2022	28/11/2022
					CT314	Electrically powered	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	28/11/2022	28/11/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	28/11/2022	28/11/2022
					CT321	Long-term surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	28/11/2022	28/11/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	28/11/2022	28/11/2022
					CT449	Generators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	28/11/2022	28/11/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	28/11/2022	28/11/2022
					CT981	Single use	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	28/11/2022	28/11/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	28/11/2022	28/11/2022
					CT983	Surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	28/11/2022	28/11/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	28/11/2022	28/11/2022
12185	2023		47263	Cardiac resynchronization therapy implantable pacemaker	CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A25	No Apparent Adverse Event	28/02/2023	28/02/2023
							F26	No Health Consequences or Impact	A25	No Apparent Adverse Event	28/02/2023	28/02/2023
					CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A25	No Apparent Adverse Event	28/02/2023	28/02/2023
							F26	No Health Consequences or Impact	A25	No Apparent Adverse Event	28/02/2023	28/02/2023
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A25	No Apparent Adverse Event	28/02/2023	28/02/2023
							F26	No Health Consequences or Impact	A25	No Apparent Adverse Event	28/02/2023	28/02/2023
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A25	No Apparent Adverse Event	28/02/2023	28/02/2023
							F26	No Health Consequences or Impact	A25	No Apparent Adverse Event	28/02/2023	28/02/2023
12186	2023		47265	Dual-chamber implantable pacemaker, rate-responsive	CT141	Electrodes/Leads and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	23/01/2023	23/01/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	23/01/2023	23/01/2023
					CT145	Cardiovascular devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	23/01/2023	23/01/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	23/01/2023	23/01/2023
					CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	23/01/2023	23/01/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	23/01/2023	23/01/2023
					CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	23/01/2023	23/01/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	23/01/2023	23/01/2023
					CT233	Surgical	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	23/01/2023	23/01/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	23/01/2023	23/01/2023
					CT2406	Implantable	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	23/01/2023	23/01/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	23/01/2023	23/01/2023
					CT2407	Active implantable	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	23/01/2023	23/01/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	23/01/2023	23/01/2023
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	23/01/2023	23/01/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	23/01/2023	23/01/2023
					CT2727	Active implant devices/ancillaries	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	23/01/2023	23/01/2023

							F1901	Additional Surgery	A24	No Apparent Device Problem	23/01/2023	23/01/2023
					CT2750	Implantable cardiac electrical stimulators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	23/01/2023	23/01/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	23/01/2023	23/01/2023
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	23/01/2023	23/01/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	23/01/2023	23/01/2023
					CT2985	Active	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	23/01/2023	23/01/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	23/01/2023	23/01/2023
					CT3001	Implantable metal	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	23/01/2023	23/01/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	23/01/2023	23/01/2023
					CT3002	Implantable synthetic polymer	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	23/01/2023	23/01/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	23/01/2023	23/01/2023
					CT3006	Implantable inorganic material	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	23/01/2023	23/01/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	23/01/2023	23/01/2023
					CT314	Electrically powered	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	23/01/2023	23/01/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	23/01/2023	23/01/2023
					CT321	Long-term surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	23/01/2023	23/01/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	23/01/2023	23/01/2023
					CT449	Generators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	23/01/2023	23/01/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	23/01/2023	23/01/2023
					CT981	Single use	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	23/01/2023	23/01/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	23/01/2023	23/01/2023
					CT983	Surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	23/01/2023	23/01/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	23/01/2023	23/01/2023
12187	2023		47265	Dual-chamber implantable pacemaker, rate-responsive	CT141	Electrodes/Leads and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	30/01/2023	30/01/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	30/01/2023	30/01/2023
					CT145	Cardiovascular devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	30/01/2023	30/01/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	30/01/2023	30/01/2023
					CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	30/01/2023	30/01/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	30/01/2023	30/01/2023
					CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	30/01/2023	30/01/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	30/01/2023	30/01/2023
					CT233	Surgical	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	30/01/2023	30/01/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	30/01/2023	30/01/2023
					CT2406	Implantable	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	30/01/2023	30/01/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	30/01/2023	30/01/2023
					CT2407	Active implantable	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	30/01/2023	30/01/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	30/01/2023	30/01/2023
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	30/01/2023	30/01/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	30/01/2023	30/01/2023
					CT2727	Active implant devices/ancillaries	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	30/01/2023	30/01/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	30/01/2023	30/01/2023
					CT2750	Implantable cardiac electrical stimulators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	30/01/2023	30/01/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	30/01/2023	30/01/2023
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	30/01/2023	30/01/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	30/01/2023	30/01/2023
					CT2985	Active	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	30/01/2023	30/01/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	30/01/2023	30/01/2023
					CT3001	Implantable metal	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	30/01/2023	30/01/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	30/01/2023	30/01/2023
					CT3002	Implantable synthetic polymer	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	30/01/2023	30/01/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	30/01/2023	30/01/2023
					CT3006	Implantable inorganic material	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	30/01/2023	30/01/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	30/01/2023	30/01/2023
					CT314	Electrically powered	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	30/01/2023	30/01/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	30/01/2023	30/01/2023
					CT321	Long-term surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	30/01/2023	30/01/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	30/01/2023	30/01/2023
					CT449	Generators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	30/01/2023	30/01/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	30/01/2023	30/01/2023
					CT981	Single use	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	30/01/2023	30/01/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	30/01/2023	30/01/2023
					CT983	Surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	30/01/2023	30/01/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	30/01/2023	30/01/2023
12188	2023		47265	Dual-chamber implantable pacemaker, rate-responsive	CT141	Electrodes/Leads and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	08/01/2023	08/01/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	08/01/2023	08/01/2023
					CT145	Cardiovascular devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	08/01/2023	08/01/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	08/01/2023	08/01/2023
					CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	08/01/2023	08/01/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	08/01/2023	08/01/2023
					CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	08/01/2023	08/01/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	08/01/2023	08/01/2023
					CT233	Surgical	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	08/01/2023	08/01/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	08/01/2023	08/01/2023
					CT2406	Implantable	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	08/01/2023	08/01/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	08/01/2023	08/01/2023
					CT2407	Active implantable	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	08/01/2023	08/01/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	08/01/2023	08/01/2023
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	08/01/2023	08/01/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	08/01/2023	08/01/2023
					CT2727	Active implant devices/ancillaries	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	08/01/2023	08/01/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	08/01/2023	08/01/2023
					CT2750	Implantable cardiac electrical stimulators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	08/01/2023	08/01/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	08/01/2023	08/01/2023
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	08/01/2023	08/01/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	08/01/2023	08/01/2023
					CT2985	Active	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	08/01/2023	08/01/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	08/01/2023	08/01/2023
					CT3001	Implantable metal	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	08/01/2023	08/01/2023

							F1901	Additional Surgery	A24	No Apparent Device Problem	08/01/2023	08/01/2023
					CT3002	Implantable synthetic polymer	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	08/01/2023	08/01/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	08/01/2023	08/01/2023
					CT3006	Implantable inorganic material	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	08/01/2023	08/01/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	08/01/2023	08/01/2023
					CT314	Electrically powered	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	08/01/2023	08/01/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	08/01/2023	08/01/2023
					CT321	Long-term surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	08/01/2023	08/01/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	08/01/2023	08/01/2023
					CT449	Generators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	08/01/2023	08/01/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	08/01/2023	08/01/2023
					CT981	Single use	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	08/01/2023	08/01/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	08/01/2023	08/01/2023
					CT983	Surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	08/01/2023	08/01/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	08/01/2023	08/01/2023
12189	2023		47265	Dual-chamber implantable pacemaker, rate-responsive	CT141	Electrodes/Leads and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	15/01/2023	15/01/2023
							F26	No Health Consequences or Impact	A24	No Apparent Device Problem	15/01/2023	15/01/2023
					CT145	Cardiovascular devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	15/01/2023	15/01/2023
							F26	No Health Consequences or Impact	A24	No Apparent Device Problem	15/01/2023	15/01/2023
					CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	15/01/2023	15/01/2023
							F26	No Health Consequences or Impact	A24	No Apparent Device Problem	15/01/2023	15/01/2023
					CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	15/01/2023	15/01/2023
							F26	No Health Consequences or Impact	A24	No Apparent Device Problem	15/01/2023	15/01/2023
					CT233	Surgical	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	15/01/2023	15/01/2023
							F26	No Health Consequences or Impact	A24	No Apparent Device Problem	15/01/2023	15/01/2023
					CT2406	Implantable	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	15/01/2023	15/01/2023
							F26	No Health Consequences or Impact	A24	No Apparent Device Problem	15/01/2023	15/01/2023
					CT2407	Active implantable	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	15/01/2023	15/01/2023
							F26	No Health Consequences or Impact	A24	No Apparent Device Problem	15/01/2023	15/01/2023
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	15/01/2023	15/01/2023
							F26	No Health Consequences or Impact	A24	No Apparent Device Problem	15/01/2023	15/01/2023
					CT2727	Active implant devices/ancillaries	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	15/01/2023	15/01/2023
							F26	No Health Consequences or Impact	A24	No Apparent Device Problem	15/01/2023	15/01/2023
					CT2750	Implantable cardiac electrical stimulators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	15/01/2023	15/01/2023
							F26	No Health Consequences or Impact	A24	No Apparent Device Problem	15/01/2023	15/01/2023
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	15/01/2023	15/01/2023
							F26	No Health Consequences or Impact	A24	No Apparent Device Problem	15/01/2023	15/01/2023
					CT2985	Active	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	15/01/2023	15/01/2023
							F26	No Health Consequences or Impact	A24	No Apparent Device Problem	15/01/2023	15/01/2023
					CT3001	Implantable metal	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	15/01/2023	15/01/2023
							F26	No Health Consequences or Impact	A24	No Apparent Device Problem	15/01/2023	15/01/2023
					CT3002	Implantable synthetic polymer	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	15/01/2023	15/01/2023
							F26	No Health Consequences or Impact	A24	No Apparent Device Problem	15/01/2023	15/01/2023
					CT3006	Implantable inorganic material	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	15/01/2023	15/01/2023
							F26	No Health Consequences or Impact	A24	No Apparent Device Problem	15/01/2023	15/01/2023
					CT314	Electrically powered	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	15/01/2023	15/01/2023
							F26	No Health Consequences or Impact	A24	No Apparent Device Problem	15/01/2023	15/01/2023
					CT321	Long-term surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	15/01/2023	15/01/2023
							F26	No Health Consequences or Impact	A24	No Apparent Device Problem	15/01/2023	15/01/2023
					CT449	Generators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	15/01/2023	15/01/2023
							F26	No Health Consequences or Impact	A24	No Apparent Device Problem	15/01/2023	15/01/2023
					CT981	Single use	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	15/01/2023	15/01/2023
							F26	No Health Consequences or Impact	A24	No Apparent Device Problem	15/01/2023	15/01/2023
					CT983	Surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	15/01/2023	15/01/2023
							F26	No Health Consequences or Impact	A24	No Apparent Device Problem	15/01/2023	15/01/2023
12190	2023	All other reportable incidents	17882	Physiologic-monitoring defibrillation system	CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070803	Failure to Power Up		
							F27	Problem Identified During Non-Clinical Procedure	A070803	Failure to Power Up		
12191	2023		47265	Dual-chamber implantable pacemaker, rate-responsive	CT141	Electrodes/Leads and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	01/10/2022	31/10/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	01/10/2022	31/10/2022
					CT145	Cardiovascular devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	01/10/2022	31/10/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	01/10/2022	31/10/2022
					CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	01/10/2022	31/10/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	01/10/2022	31/10/2022
					CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	01/10/2022	31/10/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	01/10/2022	31/10/2022
					CT233	Surgical	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	01/10/2022	31/10/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	01/10/2022	31/10/2022
					CT2406	Implantable	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	01/10/2022	31/10/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	01/10/2022	31/10/2022
					CT2407	Active implantable	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	01/10/2022	31/10/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	01/10/2022	31/10/2022
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	01/10/2022	31/10/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	01/10/2022	31/10/2022
					CT2727	Active implant devices/ancillaries	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	01/10/2022	31/10/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	01/10/2022	31/10/2022
					CT2750	Implantable cardiac electrical stimulators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	01/10/2022	31/10/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	01/10/2022	31/10/2022
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	01/10/2022	31/10/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	01/10/2022	31/10/2022
					CT2985	Active	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	01/10/2022	31/10/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	01/10/2022	31/10/2022
					CT3001	Implantable metal	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	01/10/2022	31/10/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	01/10/2022	31/10/2022
					CT3002	Implantable synthetic polymer	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	01/10/2022	31/10/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	01/10/2022	31/10/2022
					CT3006	Implantable inorganic material	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	01/10/2022	31/10/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	01/10/2022	31/10/2022
					CT314	Electrically powered	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	01/10/2022	31/10/2022

							F1901	Additional Surgery	A24	No Apparent Device Problem	01/10/2022	31/10/2022
					CT321	Long-term surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	01/10/2022	31/10/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	01/10/2022	31/10/2022
					CT449	Generators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	01/10/2022	31/10/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	01/10/2022	31/10/2022
					CT981	Single use	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	01/10/2022	31/10/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	01/10/2022	31/10/2022
					CT983	Surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	01/10/2022	31/10/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	01/10/2022	31/10/2022
12192	2023		47265	Dual-chamber implantable pacemaker, rate-responsive	CT141	Electrodes/Leads and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	01/10/2022	31/10/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	01/10/2022	31/10/2022
					CT145	Cardiovascular devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	01/10/2022	31/10/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	01/10/2022	31/10/2022
					CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	01/10/2022	31/10/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	01/10/2022	31/10/2022
					CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	01/10/2022	31/10/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	01/10/2022	31/10/2022
					CT233	Surgical	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	01/10/2022	31/10/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	01/10/2022	31/10/2022
					CT2406	Implantable	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	01/10/2022	31/10/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	01/10/2022	31/10/2022
					CT2407	Active implantable	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	01/10/2022	31/10/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	01/10/2022	31/10/2022
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	01/10/2022	31/10/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	01/10/2022	31/10/2022
					CT2727	Active implant devices/ancillaries	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	01/10/2022	31/10/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	01/10/2022	31/10/2022
					CT2750	Implantable cardiac electrical stimulators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	01/10/2022	31/10/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	01/10/2022	31/10/2022
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	01/10/2022	31/10/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	01/10/2022	31/10/2022
					CT2985	Active	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	01/10/2022	31/10/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	01/10/2022	31/10/2022
					CT3001	Implantable metal	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	01/10/2022	31/10/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	01/10/2022	31/10/2022
					CT3002	Implantable synthetic polymer	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	01/10/2022	31/10/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	01/10/2022	31/10/2022
					CT3006	Implantable inorganic material	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	01/10/2022	31/10/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	01/10/2022	31/10/2022
					CT314	Electrically powered	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	01/10/2022	31/10/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	01/10/2022	31/10/2022
					CT321	Long-term surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	01/10/2022	31/10/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	01/10/2022	31/10/2022
					CT449	Generators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	01/10/2022	31/10/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	01/10/2022	31/10/2022
					CT981	Single use	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	01/10/2022	31/10/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	01/10/2022	31/10/2022
					CT983	Surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	01/10/2022	31/10/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	01/10/2022	31/10/2022
12193	2023	All other reportable incidents	47265	Dual-chamber implantable pacemaker, rate-responsive	CT141	Electrodes/Leads and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	27/01/2023	27/01/2023
									A071102	Failure to Interrogate	27/01/2023	27/01/2023
									A071205	Pacemaker Found in Back-Up Mode	27/01/2023	27/01/2023
							F1901	Additional Surgery	A070504	Premature Discharge of Battery	27/01/2023	27/01/2023
									A071102	Failure to Interrogate	27/01/2023	27/01/2023
									A071205	Pacemaker Found in Back-Up Mode	27/01/2023	27/01/2023
					CT145	Cardiovascular devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	27/01/2023	27/01/2023
									A071102	Failure to Interrogate	27/01/2023	27/01/2023
									A071205	Pacemaker Found in Back-Up Mode	27/01/2023	27/01/2023
							F1901	Additional Surgery	A070504	Premature Discharge of Battery	27/01/2023	27/01/2023
									A071102	Failure to Interrogate	27/01/2023	27/01/2023
									A071205	Pacemaker Found in Back-Up Mode	27/01/2023	27/01/2023
					CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	27/01/2023	27/01/2023
									A071102	Failure to Interrogate	27/01/2023	27/01/2023
									A071205	Pacemaker Found in Back-Up Mode	27/01/2023	27/01/2023
							F1901	Additional Surgery	A070504	Premature Discharge of Battery	27/01/2023	27/01/2023
									A071102	Failure to Interrogate	27/01/2023	27/01/2023
									A071205	Pacemaker Found in Back-Up Mode	27/01/2023	27/01/2023
					CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	27/01/2023	27/01/2023
									A071102	Failure to Interrogate	27/01/2023	27/01/2023
									A071205	Pacemaker Found in Back-Up Mode	27/01/2023	27/01/2023
							F1901	Additional Surgery	A070504	Premature Discharge of Battery	27/01/2023	27/01/2023
									A071102	Failure to Interrogate	27/01/2023	27/01/2023
									A071205	Pacemaker Found in Back-Up Mode	27/01/2023	27/01/2023
					CT233	Surgical	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	27/01/2023	27/01/2023
									A071102	Failure to Interrogate	27/01/2023	27/01/2023
									A071205	Pacemaker Found in Back-Up Mode	27/01/2023	27/01/2023
							F1901	Additional Surgery	A070504	Premature Discharge of Battery	27/01/2023	27/01/2023
									A071102	Failure to Interrogate	27/01/2023	27/01/2023
									A071205	Pacemaker Found in Back-Up Mode	27/01/2023	27/01/2023
					CT2406	Implantable	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	27/01/2023	27/01/2023
									A071102	Failure to Interrogate	27/01/2023	27/01/2023
									A071205	Pacemaker Found in Back-Up Mode	27/01/2023	27/01/2023
							F1901	Additional Surgery	A070504	Premature Discharge of Battery	27/01/2023	27/01/2023
									A071102	Failure to Interrogate	27/01/2023	27/01/2023
									A071205	Pacemaker Found in Back-Up Mode	27/01/2023	27/01/2023
					CT2407	Active implantable	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	27/01/2023	27/01/2023
									A071102	Failure to Interrogate	27/01/2023	27/01/2023
									A071205	Pacemaker Found in Back-Up Mode	27/01/2023	27/01/2023

							F1901	Additional Surgery	A070504	Premature Discharge of Battery	27/01/2023	27/01/2023
									A071102	Failure to Interrogate	27/01/2023	27/01/2023
									A071205	Pacemaker Found in Back-Up Mode	27/01/2023	27/01/2023
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	27/01/2023	27/01/2023
									A071102	Failure to Interrogate	27/01/2023	27/01/2023
									A071205	Pacemaker Found in Back-Up Mode	27/01/2023	27/01/2023
							F1901	Additional Surgery	A070504	Premature Discharge of Battery	27/01/2023	27/01/2023
									A071102	Failure to Interrogate	27/01/2023	27/01/2023
									A071205	Pacemaker Found in Back-Up Mode	27/01/2023	27/01/2023
					CT2727	Active implant devices/ancillaries	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	27/01/2023	27/01/2023
									A071102	Failure to Interrogate	27/01/2023	27/01/2023
									A071205	Pacemaker Found in Back-Up Mode	27/01/2023	27/01/2023
							F1901	Additional Surgery	A070504	Premature Discharge of Battery	27/01/2023	27/01/2023
									A071102	Failure to Interrogate	27/01/2023	27/01/2023
									A071205	Pacemaker Found in Back-Up Mode	27/01/2023	27/01/2023
					CT2750	Implantable cardiac electrical stimulators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	27/01/2023	27/01/2023
									A071102	Failure to Interrogate	27/01/2023	27/01/2023
									A071205	Pacemaker Found in Back-Up Mode	27/01/2023	27/01/2023
							F1901	Additional Surgery	A070504	Premature Discharge of Battery	27/01/2023	27/01/2023
									A071102	Failure to Interrogate	27/01/2023	27/01/2023
									A071205	Pacemaker Found in Back-Up Mode	27/01/2023	27/01/2023
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	27/01/2023	27/01/2023
									A071102	Failure to Interrogate	27/01/2023	27/01/2023
									A071205	Pacemaker Found in Back-Up Mode	27/01/2023	27/01/2023
							F1901	Additional Surgery	A070504	Premature Discharge of Battery	27/01/2023	27/01/2023
									A071102	Failure to Interrogate	27/01/2023	27/01/2023
									A071205	Pacemaker Found in Back-Up Mode	27/01/2023	27/01/2023
					CT2985	Active	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	27/01/2023	27/01/2023
									A071102	Failure to Interrogate	27/01/2023	27/01/2023
									A071205	Pacemaker Found in Back-Up Mode	27/01/2023	27/01/2023
							F1901	Additional Surgery	A070504	Premature Discharge of Battery	27/01/2023	27/01/2023
									A071102	Failure to Interrogate	27/01/2023	27/01/2023
									A071205	Pacemaker Found in Back-Up Mode	27/01/2023	27/01/2023
					CT3001	Implantable metal	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	27/01/2023	27/01/2023
									A071102	Failure to Interrogate	27/01/2023	27/01/2023
									A071205	Pacemaker Found in Back-Up Mode	27/01/2023	27/01/2023
							F1901	Additional Surgery	A070504	Premature Discharge of Battery	27/01/2023	27/01/2023
									A071102	Failure to Interrogate	27/01/2023	27/01/2023
									A071205	Pacemaker Found in Back-Up Mode	27/01/2023	27/01/2023
					CT3002	Implantable synthetic polymer	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	27/01/2023	27/01/2023
									A071102	Failure to Interrogate	27/01/2023	27/01/2023
									A071205	Pacemaker Found in Back-Up Mode	27/01/2023	27/01/2023
							F1901	Additional Surgery	A070504	Premature Discharge of Battery	27/01/2023	27/01/2023
									A071102	Failure to Interrogate	27/01/2023	27/01/2023
									A071205	Pacemaker Found in Back-Up Mode	27/01/2023	27/01/2023
					CT3006	Implantable inorganic material	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	27/01/2023	27/01/2023
									A071102	Failure to Interrogate	27/01/2023	27/01/2023
									A071205	Pacemaker Found in Back-Up Mode	27/01/2023	27/01/2023
							F1901	Additional Surgery	A070504	Premature Discharge of Battery	27/01/2023	27/01/2023
									A071102	Failure to Interrogate	27/01/2023	27/01/2023
									A071205	Pacemaker Found in Back-Up Mode	27/01/2023	27/01/2023
					CT314	Electrically powered	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	27/01/2023	27/01/2023
									A071102	Failure to Interrogate	27/01/2023	27/01/2023
									A071205	Pacemaker Found in Back-Up Mode	27/01/2023	27/01/2023
							F1901	Additional Surgery	A070504	Premature Discharge of Battery	27/01/2023	27/01/2023
									A071102	Failure to Interrogate	27/01/2023	27/01/2023
									A071205	Pacemaker Found in Back-Up Mode	27/01/2023	27/01/2023
					CT321	Long-term surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	27/01/2023	27/01/2023
									A071102	Failure to Interrogate	27/01/2023	27/01/2023
									A071205	Pacemaker Found in Back-Up Mode	27/01/2023	27/01/2023
							F1901	Additional Surgery	A070504	Premature Discharge of Battery	27/01/2023	27/01/2023
									A071102	Failure to Interrogate	27/01/2023	27/01/2023
									A071205	Pacemaker Found in Back-Up Mode	27/01/2023	27/01/2023
					CT449	Generators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	27/01/2023	27/01/2023
									A071102	Failure to Interrogate	27/01/2023	27/01/2023
									A071205	Pacemaker Found in Back-Up Mode	27/01/2023	27/01/2023
							F1901	Additional Surgery	A070504	Premature Discharge of Battery	27/01/2023	27/01/2023
									A071102	Failure to Interrogate	27/01/2023	27/01/2023
									A071205	Pacemaker Found in Back-Up Mode	27/01/2023	27/01/2023
					CT981	Single use	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	27/01/2023	27/01/2023
									A071102	Failure to Interrogate	27/01/2023	27/01/2023
									A071205	Pacemaker Found in Back-Up Mode	27/01/2023	27/01/2023
							F1901	Additional Surgery	A070504	Premature Discharge of Battery	27/01/2023	27/01/2023
									A071102	Failure to Interrogate	27/01/2023	27/01/2023
									A071205	Pacemaker Found in Back-Up Mode	27/01/2023	27/01/2023
					CT983	Surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	27/01/2023	27/01/2023
									A071102	Failure to Interrogate	27/01/2023	27/01/2023
									A071205	Pacemaker Found in Back-Up Mode	27/01/2023	27/01/2023
							F1901	Additional Surgery	A070504	Premature Discharge of Battery	27/01/2023	27/01/2023
									A071102	Failure to Interrogate	27/01/2023	27/01/2023
									A071205	Pacemaker Found in Back-Up Mode	27/01/2023	27/01/2023
12194	2023	All other reportable incidents	17882	Physiologic-monitoring defibrillation system	CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A07	Electrical /Electronic Property Problem		
							F27	Problem Identified During Non-Clinical Procedure	A07	Electrical /Electronic Property Problem		
12195	2023	All other reportable incidents	17882	Physiologic-monitoring defibrillation system	CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A071301	Failure to Deliver Shock/Stimulation		
							F27	Problem Identified During Non-Clinical Procedure	A071301	Failure to Deliver Shock/Stimulation		
12196	2023	All other reportable incidents	47267	Single-chamber implantable pacemaker, rate-responsive	CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A0705	Battery Problem		

								A070504	Premature Discharge of Battery			
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
12197	2023	All other reportable incidents	47265	Dual-chamber implantable pacemaker, rate-responsive	CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	05/01/2023	05/01/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	05/01/2023	05/01/2023
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	05/01/2023	05/01/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	05/01/2023	05/01/2023
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	05/01/2023	05/01/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	05/01/2023	05/01/2023
12198	2023	All other reportable incidents	47265	Dual-chamber implantable pacemaker, rate-responsive	CT141	Electrodes/Leads and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	29/11/2022	29/11/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	29/11/2022	29/11/2022
					CT145	Cardiovascular devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	29/11/2022	29/11/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	29/11/2022	29/11/2022
					CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	29/11/2022	29/11/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	29/11/2022	29/11/2022
					CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	29/11/2022	29/11/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	29/11/2022	29/11/2022
					CT233	Surgical	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	29/11/2022	29/11/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	29/11/2022	29/11/2022
					CT2406	Implantable	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	29/11/2022	29/11/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	29/11/2022	29/11/2022
					CT2407	Active implantable	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	29/11/2022	29/11/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	29/11/2022	29/11/2022
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	29/11/2022	29/11/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	29/11/2022	29/11/2022
					CT2727	Active implant devices/ancillaries	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	29/11/2022	29/11/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	29/11/2022	29/11/2022
					CT2750	Implantable cardiac electrical stimulators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	29/11/2022	29/11/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	29/11/2022	29/11/2022
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	29/11/2022	29/11/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	29/11/2022	29/11/2022
					CT2985	Active	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	29/11/2022	29/11/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	29/11/2022	29/11/2022
					CT3001	Implantable metal	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	29/11/2022	29/11/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	29/11/2022	29/11/2022
					CT3002	Implantable synthetic polymer	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	29/11/2022	29/11/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	29/11/2022	29/11/2022
					CT3006	Implantable inorganic material	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	29/11/2022	29/11/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	29/11/2022	29/11/2022
					CT314	Electrically powered	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	29/11/2022	29/11/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	29/11/2022	29/11/2022
					CT321	Long-term surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	29/11/2022	29/11/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	29/11/2022	29/11/2022
					CT449	Generators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	29/11/2022	29/11/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	29/11/2022	29/11/2022
					CT981	Single use	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	29/11/2022	29/11/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	29/11/2022	29/11/2022
					CT983	Surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	29/11/2022	29/11/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	29/11/2022	29/11/2022
12199	2023	All other reportable incidents	47265	Dual-chamber implantable pacemaker, rate-responsive	CT141	Electrodes/Leads and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem	12/11/2022	12/11/2022
									A070504	Premature Discharge of Battery	12/11/2022	12/11/2022
							F08	Hospitalization or Prolonged Hospitalization	A0705	Battery Problem	12/11/2022	12/11/2022
									A070504	Premature Discharge of Battery	12/11/2022	12/11/2022
							F1905	Device Revision or Replacement	A0705	Battery Problem	12/11/2022	12/11/2022
									A070504	Premature Discharge of Battery	12/11/2022	12/11/2022
					CT145	Cardiovascular devices	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem	12/11/2022	12/11/2022
									A070504	Premature Discharge of Battery	12/11/2022	12/11/2022
							F08	Hospitalization or Prolonged Hospitalization	A0705	Battery Problem	12/11/2022	12/11/2022
									A070504	Premature Discharge of Battery	12/11/2022	12/11/2022
							F1905	Device Revision or Replacement	A0705	Battery Problem	12/11/2022	12/11/2022
									A070504	Premature Discharge of Battery	12/11/2022	12/11/2022
					CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem	12/11/2022	12/11/2022
									A070504	Premature Discharge of Battery	12/11/2022	12/11/2022
							F08	Hospitalization or Prolonged Hospitalization	A0705	Battery Problem	12/11/2022	12/11/2022
									A070504	Premature Discharge of Battery	12/11/2022	12/11/2022
							F1905	Device Revision or Replacement	A0705	Battery Problem	12/11/2022	12/11/2022
									A070504	Premature Discharge of Battery	12/11/2022	12/11/2022
					CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem	12/11/2022	12/11/2022
									A070504	Premature Discharge of Battery	12/11/2022	12/11/2022
							F08	Hospitalization or Prolonged Hospitalization	A0705	Battery Problem	12/11/2022	12/11/2022
									A070504	Premature Discharge of Battery	12/11/2022	12/11/2022
							F1905	Device Revision or Replacement	A0705	Battery Problem	12/11/2022	12/11/2022
									A070504	Premature Discharge of Battery	12/11/2022	12/11/2022
					CT233	Surgical	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem	12/11/2022	12/11/2022
									A070504	Premature Discharge of Battery	12/11/2022	12/11/2022
							F08	Hospitalization or Prolonged Hospitalization	A0705	Battery Problem	12/11/2022	12/11/2022
									A070504	Premature Discharge of Battery	12/11/2022	12/11/2022
							F1905	Device Revision or Replacement	A0705	Battery Problem	12/11/2022	12/11/2022
									A070504	Premature Discharge of Battery	12/11/2022	12/11/2022
					CT2406	Implantable	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem	12/11/2022	12/11/2022
									A070504	Premature Discharge of Battery	12/11/2022	12/11/2022
							F08	Hospitalization or Prolonged Hospitalization	A0705	Battery Problem	12/11/2022	12/11/2022

									A070504	Premature Discharge of Battery	12/11/2022	12/11/2022
							F1905	Device Revision or Replacement	A0705	Battery Problem	12/11/2022	12/11/2022
									A070504	Premature Discharge of Battery	12/11/2022	12/11/2022
					CT2407	Active implantable	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem	12/11/2022	12/11/2022
									A070504	Premature Discharge of Battery	12/11/2022	12/11/2022
							F08	Hospitalization or Prolonged Hospitalization	A0705	Battery Problem	12/11/2022	12/11/2022
									A070504	Premature Discharge of Battery	12/11/2022	12/11/2022
							F1905	Device Revision or Replacement	A0705	Battery Problem	12/11/2022	12/11/2022
									A070504	Premature Discharge of Battery	12/11/2022	12/11/2022
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem	12/11/2022	12/11/2022
									A070504	Premature Discharge of Battery	12/11/2022	12/11/2022
							F08	Hospitalization or Prolonged Hospitalization	A0705	Battery Problem	12/11/2022	12/11/2022
									A070504	Premature Discharge of Battery	12/11/2022	12/11/2022
							F1905	Device Revision or Replacement	A0705	Battery Problem	12/11/2022	12/11/2022
									A070504	Premature Discharge of Battery	12/11/2022	12/11/2022
					CT2727	Active implant devices/ancillaries	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem	12/11/2022	12/11/2022
									A070504	Premature Discharge of Battery	12/11/2022	12/11/2022
							F08	Hospitalization or Prolonged Hospitalization	A0705	Battery Problem	12/11/2022	12/11/2022
									A070504	Premature Discharge of Battery	12/11/2022	12/11/2022
							F1905	Device Revision or Replacement	A0705	Battery Problem	12/11/2022	12/11/2022
									A070504	Premature Discharge of Battery	12/11/2022	12/11/2022
					CT2750	Implantable cardiac electrical stimulators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem	12/11/2022	12/11/2022
									A070504	Premature Discharge of Battery	12/11/2022	12/11/2022
							F08	Hospitalization or Prolonged Hospitalization	A0705	Battery Problem	12/11/2022	12/11/2022
									A070504	Premature Discharge of Battery	12/11/2022	12/11/2022
							F1905	Device Revision or Replacement	A0705	Battery Problem	12/11/2022	12/11/2022
									A070504	Premature Discharge of Battery	12/11/2022	12/11/2022
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem	12/11/2022	12/11/2022
									A070504	Premature Discharge of Battery	12/11/2022	12/11/2022
							F08	Hospitalization or Prolonged Hospitalization	A0705	Battery Problem	12/11/2022	12/11/2022
									A070504	Premature Discharge of Battery	12/11/2022	12/11/2022
							F1905	Device Revision or Replacement	A0705	Battery Problem	12/11/2022	12/11/2022
									A070504	Premature Discharge of Battery	12/11/2022	12/11/2022
					CT2985	Active	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem	12/11/2022	12/11/2022
									A070504	Premature Discharge of Battery	12/11/2022	12/11/2022
							F08	Hospitalization or Prolonged Hospitalization	A0705	Battery Problem	12/11/2022	12/11/2022
									A070504	Premature Discharge of Battery	12/11/2022	12/11/2022
							F1905	Device Revision or Replacement	A0705	Battery Problem	12/11/2022	12/11/2022
									A070504	Premature Discharge of Battery	12/11/2022	12/11/2022
					CT3001	Implantable metal	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem	12/11/2022	12/11/2022
									A070504	Premature Discharge of Battery	12/11/2022	12/11/2022
							F08	Hospitalization or Prolonged Hospitalization	A0705	Battery Problem	12/11/2022	12/11/2022
									A070504	Premature Discharge of Battery	12/11/2022	12/11/2022
							F1905	Device Revision or Replacement	A0705	Battery Problem	12/11/2022	12/11/2022
									A070504	Premature Discharge of Battery	12/11/2022	12/11/2022
					CT3002	Implantable synthetic polymer	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem	12/11/2022	12/11/2022
									A070504	Premature Discharge of Battery	12/11/2022	12/11/2022
							F08	Hospitalization or Prolonged Hospitalization	A0705	Battery Problem	12/11/2022	12/11/2022
									A070504	Premature Discharge of Battery	12/11/2022	12/11/2022
							F1905	Device Revision or Replacement	A0705	Battery Problem	12/11/2022	12/11/2022
									A070504	Premature Discharge of Battery	12/11/2022	12/11/2022
					CT3006	Implantable inorganic material	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem	12/11/2022	12/11/2022
									A070504	Premature Discharge of Battery	12/11/2022	12/11/2022
							F08	Hospitalization or Prolonged Hospitalization	A0705	Battery Problem	12/11/2022	12/11/2022
									A070504	Premature Discharge of Battery	12/11/2022	12/11/2022
							F1905	Device Revision or Replacement	A0705	Battery Problem	12/11/2022	12/11/2022
									A070504	Premature Discharge of Battery	12/11/2022	12/11/2022
					CT314	Electrically powered	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem	12/11/2022	12/11/2022
									A070504	Premature Discharge of Battery	12/11/2022	12/11/2022
							F08	Hospitalization or Prolonged Hospitalization	A0705	Battery Problem	12/11/2022	12/11/2022
									A070504	Premature Discharge of Battery	12/11/2022	12/11/2022
							F1905	Device Revision or Replacement	A0705	Battery Problem	12/11/2022	12/11/2022
									A070504	Premature Discharge of Battery	12/11/2022	12/11/2022
					CT321	Long-term surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem	12/11/2022	12/11/2022
									A070504	Premature Discharge of Battery	12/11/2022	12/11/2022
							F08	Hospitalization or Prolonged Hospitalization	A0705	Battery Problem	12/11/2022	12/11/2022
									A070504	Premature Discharge of Battery	12/11/2022	12/11/2022
							F1905	Device Revision or Replacement	A0705	Battery Problem	12/11/2022	12/11/2022
									A070504	Premature Discharge of Battery	12/11/2022	12/11/2022
					CT449	Generators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem	12/11/2022	12/11/2022
									A070504	Premature Discharge of Battery	12/11/2022	12/11/2022
							F08	Hospitalization or Prolonged Hospitalization	A0705	Battery Problem	12/11/2022	12/11/2022
									A070504	Premature Discharge of Battery	12/11/2022	12/11/2022
							F1905	Device Revision or Replacement	A0705	Battery Problem	12/11/2022	12/11/2022
									A070504	Premature Discharge of Battery	12/11/2022	12/11/2022
					CT981	Single use	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem	12/11/2022	12/11/2022
									A070504	Premature Discharge of Battery	12/11/2022	12/11/2022
							F08	Hospitalization or Prolonged Hospitalization	A0705	Battery Problem	12/11/2022	12/11/2022
									A070504	Premature Discharge of Battery	12/11/2022	12/11/2022
							F1905	Device Revision or Replacement	A0705	Battery Problem	12/11/2022	12/11/2022
									A070504	Premature Discharge of Battery	12/11/2022	12/11/2022
					CT983	Surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem	12/11/2022	12/11/2022
									A070504	Premature Discharge of Battery	12/11/2022	12/11/2022
							F08	Hospitalization or Prolonged Hospitalization	A0705	Battery Problem	12/11/2022	12/11/2022
									A070504	Premature Discharge of Battery	12/11/2022	12/11/2022
							F1905	Device Revision or Replacement	A0705	Battery Problem	12/11/2022	12/11/2022
									A070504	Premature Discharge of Battery	12/11/2022	12/11/2022
12200	2023	All other reportable incidents	47265	Dual-chamber implantable pacemaker, rate-	CT141	Electrodes/Leads and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem		

				responsive			F1901	Additional Surgery	A24	No Apparent Device Problem		
					CT145	Cardiovascular devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem		
							F1901	Additional Surgery	A24	No Apparent Device Problem		
					CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem		
							F1901	Additional Surgery	A24	No Apparent Device Problem		
					CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem		
							F1901	Additional Surgery	A24	No Apparent Device Problem		
					CT233	Surgical	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem		
							F1901	Additional Surgery	A24	No Apparent Device Problem		
					CT2406	Implantable	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem		
							F1901	Additional Surgery	A24	No Apparent Device Problem		
					CT2407	Active implantable	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem		
							F1901	Additional Surgery	A24	No Apparent Device Problem		
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem		
							F1901	Additional Surgery	A24	No Apparent Device Problem		
					CT2727	Active implant devices/ancillaries	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem		
							F1901	Additional Surgery	A24	No Apparent Device Problem		
					CT2750	Implantable cardiac electrical stimulators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem		
							F1901	Additional Surgery	A24	No Apparent Device Problem		
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem		
							F1901	Additional Surgery	A24	No Apparent Device Problem		
					CT2985	Active	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem		
							F1901	Additional Surgery	A24	No Apparent Device Problem		
					CT3001	Implantable metal	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem		
							F1901	Additional Surgery	A24	No Apparent Device Problem		
					CT3002	Implantable synthetic polymer	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem		
							F1901	Additional Surgery	A24	No Apparent Device Problem		
					CT3006	Implantable inorganic material	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem		
							F1901	Additional Surgery	A24	No Apparent Device Problem		
					CT314	Electrically powered	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem		
							F1901	Additional Surgery	A24	No Apparent Device Problem		
					CT321	Long-term surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem		
							F1901	Additional Surgery	A24	No Apparent Device Problem		
					CT449	Generators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem		
							F1901	Additional Surgery	A24	No Apparent Device Problem		
					CT981	Single use	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem		
							F1901	Additional Surgery	A24	No Apparent Device Problem		
					CT983	Surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem		
							F1901	Additional Surgery	A24	No Apparent Device Problem		
12201	2023	All other reportable incidents	47270	Cardiac resynchronization therapy implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E2312	Fatigue	A160503	Inappropriate or Unexpected Reset	10/02/2023	10/02/2023
							F1901	Additional Surgery	A160503	Inappropriate or Unexpected Reset	10/02/2023	10/02/2023
12202	2023		47265	Dual-chamber implantable pacemaker, rate-responsive	CT141	Electrodes/Leads and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	01/10/2022	31/10/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	01/10/2022	31/10/2022
					CT145	Cardiovascular devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	01/10/2022	31/10/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	01/10/2022	31/10/2022
					CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	01/10/2022	31/10/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	01/10/2022	31/10/2022
					CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	01/10/2022	31/10/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	01/10/2022	31/10/2022
					CT233	Surgical	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	01/10/2022	31/10/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	01/10/2022	31/10/2022
					CT2406	Implantable	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	01/10/2022	31/10/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	01/10/2022	31/10/2022
					CT2407	Active implantable	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	01/10/2022	31/10/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	01/10/2022	31/10/2022
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	01/10/2022	31/10/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	01/10/2022	31/10/2022
					CT2727	Active implant devices/ancillaries	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	01/10/2022	31/10/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	01/10/2022	31/10/2022
					CT2750	Implantable cardiac electrical stimulators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	01/10/2022	31/10/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	01/10/2022	31/10/2022
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	01/10/2022	31/10/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	01/10/2022	31/10/2022
					CT2985	Active	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	01/10/2022	31/10/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	01/10/2022	31/10/2022
					CT3001	Implantable metal	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	01/10/2022	31/10/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	01/10/2022	31/10/2022
					CT3002	Implantable synthetic polymer	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	01/10/2022	31/10/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	01/10/2022	31/10/2022
					CT3006	Implantable inorganic material	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	01/10/2022	31/10/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	01/10/2022	31/10/2022
					CT314	Electrically powered	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	01/10/2022	31/10/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	01/10/2022	31/10/2022
					CT321	Long-term surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	01/10/2022	31/10/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	01/10/2022	31/10/2022
					CT449	Generators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	01/10/2022	31/10/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	01/10/2022	31/10/2022
					CT981	Single use	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	01/10/2022	31/10/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	01/10/2022	31/10/2022
					CT983	Surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	01/10/2022	31/10/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	01/10/2022	31/10/2022
12203	2023	All other reportable incidents	47263	Cardiac resynchronization therapy implantable pacemaker	CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		

							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
12204	2023		47265	Dual-chamber implantable pacemaker, rate-responsive	CT141	Electrodes/Leads and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem		
							F1901	Additional Surgery	A24	No Apparent Device Problem		
					CT145	Cardiovascular devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem		
							F1901	Additional Surgery	A24	No Apparent Device Problem		
					CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem		
							F1901	Additional Surgery	A24	No Apparent Device Problem		
					CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem		
							F1901	Additional Surgery	A24	No Apparent Device Problem		
					CT233	Surgical	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem		
							F1901	Additional Surgery	A24	No Apparent Device Problem		
					CT2406	Implantable	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem		
							F1901	Additional Surgery	A24	No Apparent Device Problem		
					CT2407	Active implantable	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem		
							F1901	Additional Surgery	A24	No Apparent Device Problem		
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem		
							F1901	Additional Surgery	A24	No Apparent Device Problem		
					CT2727	Active implant devices/ancillaries	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem		
							F1901	Additional Surgery	A24	No Apparent Device Problem		
					CT2750	Implantable cardiac electrical stimulators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem		
							F1901	Additional Surgery	A24	No Apparent Device Problem		
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem		
							F1901	Additional Surgery	A24	No Apparent Device Problem		
					CT2985	Active	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem		
							F1901	Additional Surgery	A24	No Apparent Device Problem		
					CT3001	Implantable metal	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem		
							F1901	Additional Surgery	A24	No Apparent Device Problem		
					CT3002	Implantable synthetic polymer	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem		
							F1901	Additional Surgery	A24	No Apparent Device Problem		
					CT3006	Implantable inorganic material	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem		
							F1901	Additional Surgery	A24	No Apparent Device Problem		
					CT314	Electrically powered	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem		
							F1901	Additional Surgery	A24	No Apparent Device Problem		
					CT321	Long-term surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem		
							F1901	Additional Surgery	A24	No Apparent Device Problem		
					CT449	Generators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem		
							F1901	Additional Surgery	A24	No Apparent Device Problem		
					CT981	Single use	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem		
							F1901	Additional Surgery	A24	No Apparent Device Problem		
					CT983	Surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem		
							F1901	Additional Surgery	A24	No Apparent Device Problem		
12205	2023		47265	Dual-chamber implantable pacemaker, rate-responsive	CT141	Electrodes/Leads and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	08/01/2023	08/01/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	08/01/2023	08/01/2023
					CT145	Cardiovascular devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	08/01/2023	08/01/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	08/01/2023	08/01/2023
					CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	08/01/2023	08/01/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	08/01/2023	08/01/2023
					CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	08/01/2023	08/01/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	08/01/2023	08/01/2023
					CT233	Surgical	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	08/01/2023	08/01/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	08/01/2023	08/01/2023
					CT2406	Implantable	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	08/01/2023	08/01/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	08/01/2023	08/01/2023
					CT2407	Active implantable	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	08/01/2023	08/01/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	08/01/2023	08/01/2023
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	08/01/2023	08/01/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	08/01/2023	08/01/2023
					CT2727	Active implant devices/ancillaries	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	08/01/2023	08/01/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	08/01/2023	08/01/2023
					CT2750	Implantable cardiac electrical stimulators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	08/01/2023	08/01/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	08/01/2023	08/01/2023
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	08/01/2023	08/01/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	08/01/2023	08/01/2023
					CT2985	Active	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	08/01/2023	08/01/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	08/01/2023	08/01/2023
					CT3001	Implantable metal	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	08/01/2023	08/01/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	08/01/2023	08/01/2023
					CT3002	Implantable synthetic polymer	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	08/01/2023	08/01/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	08/01/2023	08/01/2023
					CT3006	Implantable inorganic material	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	08/01/2023	08/01/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	08/01/2023	08/01/2023
					CT314	Electrically powered	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	08/01/2023	08/01/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	08/01/2023	08/01/2023
					CT321	Long-term surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	08/01/2023	08/01/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	08/01/2023	08/01/2023
					CT449	Generators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	08/01/2023	08/01/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	08/01/2023	08/01/2023
					CT981	Single use	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	08/01/2023	08/01/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	08/01/2023	08/01/2023
					CT983	Surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	08/01/2023	08/01/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	08/01/2023	08/01/2023
12206	2023	All other reportable incidents	47265	Dual-chamber implantable pacemaker, rate-responsive	CT141	Electrodes/Leads and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	20/01/2023	20/01/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	20/01/2023	20/01/2023
					CT145	Cardiovascular devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	20/01/2023	20/01/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	20/01/2023	20/01/2023
					CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	20/01/2023	20/01/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	20/01/2023	20/01/2023
					CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	20/01/2023	20/01/2023

							F1901	Additional Surgery	A24	No Apparent Device Problem	20/01/2023	20/01/2023
					CT233	Surgical	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	20/01/2023	20/01/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	20/01/2023	20/01/2023
					CT2406	Implantable	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	20/01/2023	20/01/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	20/01/2023	20/01/2023
					CT2407	Active implantable	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	20/01/2023	20/01/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	20/01/2023	20/01/2023
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	20/01/2023	20/01/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	20/01/2023	20/01/2023
					CT2727	Active implant devices/ancillaries	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	20/01/2023	20/01/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	20/01/2023	20/01/2023
					CT2750	Implantable cardiac electrical stimulators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	20/01/2023	20/01/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	20/01/2023	20/01/2023
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	20/01/2023	20/01/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	20/01/2023	20/01/2023
					CT2985	Active	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	20/01/2023	20/01/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	20/01/2023	20/01/2023
					CT3001	Implantable metal	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	20/01/2023	20/01/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	20/01/2023	20/01/2023
					CT3002	Implantable synthetic polymer	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	20/01/2023	20/01/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	20/01/2023	20/01/2023
					CT3006	Implantable inorganic material	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	20/01/2023	20/01/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	20/01/2023	20/01/2023
					CT314	Electrically powered	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	20/01/2023	20/01/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	20/01/2023	20/01/2023
					CT321	Long-term surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	20/01/2023	20/01/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	20/01/2023	20/01/2023
					CT449	Generators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	20/01/2023	20/01/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	20/01/2023	20/01/2023
					CT981	Single use	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	20/01/2023	20/01/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	20/01/2023	20/01/2023
					CT983	Surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	20/01/2023	20/01/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	20/01/2023	20/01/2023
12207	2023		47265	Dual-chamber implantable pacemaker, rate-responsive	CT141	Electrodes/Leads and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	15/01/2023	15/01/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	15/01/2023	15/01/2023
					CT145	Cardiovascular devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	15/01/2023	15/01/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	15/01/2023	15/01/2023
					CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	15/01/2023	15/01/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	15/01/2023	15/01/2023
					CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	15/01/2023	15/01/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	15/01/2023	15/01/2023
					CT233	Surgical	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	15/01/2023	15/01/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	15/01/2023	15/01/2023
					CT2406	Implantable	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	15/01/2023	15/01/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	15/01/2023	15/01/2023
					CT2407	Active implantable	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	15/01/2023	15/01/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	15/01/2023	15/01/2023
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	15/01/2023	15/01/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	15/01/2023	15/01/2023
					CT2727	Active implant devices/ancillaries	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	15/01/2023	15/01/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	15/01/2023	15/01/2023
					CT2750	Implantable cardiac electrical stimulators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	15/01/2023	15/01/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	15/01/2023	15/01/2023
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	15/01/2023	15/01/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	15/01/2023	15/01/2023
					CT2985	Active	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	15/01/2023	15/01/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	15/01/2023	15/01/2023
					CT3001	Implantable metal	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	15/01/2023	15/01/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	15/01/2023	15/01/2023
					CT3002	Implantable synthetic polymer	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	15/01/2023	15/01/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	15/01/2023	15/01/2023
					CT3006	Implantable inorganic material	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	15/01/2023	15/01/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	15/01/2023	15/01/2023
					CT314	Electrically powered	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	15/01/2023	15/01/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	15/01/2023	15/01/2023
					CT321	Long-term surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	15/01/2023	15/01/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	15/01/2023	15/01/2023
					CT449	Generators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	15/01/2023	15/01/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	15/01/2023	15/01/2023
					CT981	Single use	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	15/01/2023	15/01/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	15/01/2023	15/01/2023
					CT983	Surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	15/01/2023	15/01/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	15/01/2023	15/01/2023
12208	2023		47265	Dual-chamber implantable pacemaker, rate-responsive	CT141	Electrodes/Leads and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	15/01/2023	15/01/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	15/01/2023	15/01/2023
					CT145	Cardiovascular devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	15/01/2023	15/01/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	15/01/2023	15/01/2023
					CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	15/01/2023	15/01/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	15/01/2023	15/01/2023
					CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	15/01/2023	15/01/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	15/01/2023	15/01/2023
					CT233	Surgical	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	15/01/2023	15/01/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	15/01/2023	15/01/2023
					CT2406	Implantable	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	15/01/2023	15/01/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	15/01/2023	15/01/2023
					CT2407	Active implantable	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	15/01/2023	15/01/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	15/01/2023	15/01/2023
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	15/01/2023	15/01/2023

						F1901	Additional Surgery	A24	No Apparent Device Problem	15/01/2023	15/01/2023
					CT2727	Active implant devices/ancillaries	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	15/01/2023	15/01/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	15/01/2023	15/01/2023
					CT2750	Implantable cardiac electrical stimulators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	15/01/2023	15/01/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	15/01/2023	15/01/2023
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	15/01/2023	15/01/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	15/01/2023	15/01/2023
					CT2985	Active	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	15/01/2023	15/01/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	15/01/2023	15/01/2023
					CT3001	Implantable metal	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	15/01/2023	15/01/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	15/01/2023	15/01/2023
					CT3002	Implantable synthetic polymer	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	15/01/2023	15/01/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	15/01/2023	15/01/2023
					CT3006	Implantable inorganic material	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	15/01/2023	15/01/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	15/01/2023	15/01/2023
					CT314	Electrically powered	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	15/01/2023	15/01/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	15/01/2023	15/01/2023
					CT321	Long-term surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	15/01/2023	15/01/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	15/01/2023	15/01/2023
					CT449	Generators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	15/01/2023	15/01/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	15/01/2023	15/01/2023
					CT981	Single use	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	15/01/2023	15/01/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	15/01/2023	15/01/2023
					CT983	Surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	15/01/2023	15/01/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	15/01/2023	15/01/2023
12209	2023		47265	Dual-chamber implantable pacemaker, rate-responsive	CT141	Electrodes/Leads and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	24/01/2023	24/01/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	24/01/2023	24/01/2023
					CT145	Cardiovascular devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	24/01/2023	24/01/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	24/01/2023	24/01/2023
					CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	24/01/2023	24/01/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	24/01/2023	24/01/2023
					CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	24/01/2023	24/01/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	24/01/2023	24/01/2023
					CT233	Surgical	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	24/01/2023	24/01/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	24/01/2023	24/01/2023
					CT2406	Implantable	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	24/01/2023	24/01/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	24/01/2023	24/01/2023
					CT2407	Active implantable	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	24/01/2023	24/01/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	24/01/2023	24/01/2023
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	24/01/2023	24/01/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	24/01/2023	24/01/2023
					CT2727	Active implant devices/ancillaries	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	24/01/2023	24/01/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	24/01/2023	24/01/2023
					CT2750	Implantable cardiac electrical stimulators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	24/01/2023	24/01/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	24/01/2023	24/01/2023
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	24/01/2023	24/01/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	24/01/2023	24/01/2023
					CT2985	Active	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	24/01/2023	24/01/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	24/01/2023	24/01/2023
					CT3001	Implantable metal	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	24/01/2023	24/01/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	24/01/2023	24/01/2023
					CT3002	Implantable synthetic polymer	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	24/01/2023	24/01/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	24/01/2023	24/01/2023
					CT3006	Implantable inorganic material	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	24/01/2023	24/01/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	24/01/2023	24/01/2023
					CT314	Electrically powered	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	24/01/2023	24/01/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	24/01/2023	24/01/2023
					CT321	Long-term surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	24/01/2023	24/01/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	24/01/2023	24/01/2023
					CT449	Generators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	24/01/2023	24/01/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	24/01/2023	24/01/2023
					CT981	Single use	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	24/01/2023	24/01/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	24/01/2023	24/01/2023
					CT983	Surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	24/01/2023	24/01/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	24/01/2023	24/01/2023
12210	2023		47265	Dual-chamber implantable pacemaker, rate-responsive	CT141	Electrodes/Leads and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	26/01/2023	26/01/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	26/01/2023	26/01/2023
					CT145	Cardiovascular devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	26/01/2023	26/01/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	26/01/2023	26/01/2023
					CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	26/01/2023	26/01/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	26/01/2023	26/01/2023
					CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	26/01/2023	26/01/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	26/01/2023	26/01/2023
					CT233	Surgical	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	26/01/2023	26/01/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	26/01/2023	26/01/2023
					CT2406	Implantable	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	26/01/2023	26/01/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	26/01/2023	26/01/2023
					CT2407	Active implantable	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	26/01/2023	26/01/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	26/01/2023	26/01/2023
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	26/01/2023	26/01/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	26/01/2023	26/01/2023
					CT2727	Active implant devices/ancillaries	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	26/01/2023	26/01/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	26/01/2023	26/01/2023
					CT2750	Implantable cardiac electrical stimulators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	26/01/2023	26/01/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	26/01/2023	26/01/2023
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	26/01/2023	26/01/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	26/01/2023	26/01/2023
					CT2985	Active	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	26/01/2023	26/01/2023

							F1901	Additional Surgery	A24	No Apparent Device Problem	26/01/2023	26/01/2023
					CT3001	Implantable metal	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	26/01/2023	26/01/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	26/01/2023	26/01/2023
					CT3002	Implantable synthetic polymer	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	26/01/2023	26/01/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	26/01/2023	26/01/2023
					CT3006	Implantable inorganic material	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	26/01/2023	26/01/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	26/01/2023	26/01/2023
					CT314	Electrically powered	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	26/01/2023	26/01/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	26/01/2023	26/01/2023
					CT321	Long-term surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	26/01/2023	26/01/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	26/01/2023	26/01/2023
					CT449	Generators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	26/01/2023	26/01/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	26/01/2023	26/01/2023
					CT981	Single use	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	26/01/2023	26/01/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	26/01/2023	26/01/2023
					CT983	Surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	26/01/2023	26/01/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	26/01/2023	26/01/2023
12211	2023		47265	Dual-chamber implantable pacemaker, rate-responsive	CT141	Electrodes/Leads and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	25/01/2023	25/01/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	25/01/2023	25/01/2023
					CT145	Cardiovascular devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	25/01/2023	25/01/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	25/01/2023	25/01/2023
					CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	25/01/2023	25/01/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	25/01/2023	25/01/2023
					CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	25/01/2023	25/01/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	25/01/2023	25/01/2023
					CT233	Surgical	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	25/01/2023	25/01/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	25/01/2023	25/01/2023
					CT2406	Implantable	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	25/01/2023	25/01/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	25/01/2023	25/01/2023
					CT2407	Active implantable	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	25/01/2023	25/01/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	25/01/2023	25/01/2023
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	25/01/2023	25/01/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	25/01/2023	25/01/2023
					CT2727	Active implant devices/ancillaries	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	25/01/2023	25/01/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	25/01/2023	25/01/2023
					CT2750	Implantable cardiac electrical stimulators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	25/01/2023	25/01/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	25/01/2023	25/01/2023
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	25/01/2023	25/01/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	25/01/2023	25/01/2023
					CT2985	Active	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	25/01/2023	25/01/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	25/01/2023	25/01/2023
					CT3001	Implantable metal	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	25/01/2023	25/01/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	25/01/2023	25/01/2023
					CT3002	Implantable synthetic polymer	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	25/01/2023	25/01/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	25/01/2023	25/01/2023
					CT3006	Implantable inorganic material	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	25/01/2023	25/01/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	25/01/2023	25/01/2023
					CT314	Electrically powered	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	25/01/2023	25/01/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	25/01/2023	25/01/2023
					CT321	Long-term surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	25/01/2023	25/01/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	25/01/2023	25/01/2023
					CT449	Generators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	25/01/2023	25/01/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	25/01/2023	25/01/2023
					CT981	Single use	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	25/01/2023	25/01/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	25/01/2023	25/01/2023
					CT983	Surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	25/01/2023	25/01/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	25/01/2023	25/01/2023
12212	2023	All other reportable incidents	12913	Pacemaker, implantable	CT145	Cardiovascular devices	E2403	No Clinical Signs, Symptoms or Conditions	A070101	Failure to Capture		
									A070102	High Capture Threshold		
									A072202	Low Impedance		
									A1301	Failure to Read Input Signal		
							F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A070101	Failure to Capture		
									A070102	High Capture Threshold		
									A072202	Low Impedance		
									A1301	Failure to Read Input Signal		
							F2203	Imaging Required	A070101	Failure to Capture		
									A070102	High Capture Threshold		
									A072202	Low Impedance		
									A1301	Failure to Read Input Signal		
					CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070101	Failure to Capture		
									A070102	High Capture Threshold		
									A072202	Low Impedance		
									A1301	Failure to Read Input Signal		
							F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A070101	Failure to Capture		
									A070102	High Capture Threshold		
									A072202	Low Impedance		
									A1301	Failure to Read Input Signal		
							F2203	Imaging Required	A070101	Failure to Capture		
									A070102	High Capture Threshold		
									A072202	Low Impedance		
									A1301	Failure to Read Input Signal		
					CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A070101	Failure to Capture		
									A070102	High Capture Threshold		
									A072202	Low Impedance		
									A1301	Failure to Read Input Signal		
							F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A070101	Failure to Capture		
									A070102	High Capture Threshold		
									A072202	Low Impedance		

									A1301	Failure to Read Input Signal		
							F2203	Imaging Required	A070101	Failure to Capture		
									A070102	High Capture Threshold		
									A072202	Low Impedance		
									A1301	Failure to Read Input Signal		
					CT2406	Implantable	E2403	No Clinical Signs, Symptoms or Conditions	A070101	Failure to Capture		
									A070102	High Capture Threshold		
									A072202	Low Impedance		
									A1301	Failure to Read Input Signal		
							F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A070101	Failure to Capture		
									A070102	High Capture Threshold		
									A072202	Low Impedance		
									A1301	Failure to Read Input Signal		
							F2203	Imaging Required	A070101	Failure to Capture		
									A070102	High Capture Threshold		
									A072202	Low Impedance		
									A1301	Failure to Read Input Signal		
					CT2407	Active implantable	E2403	No Clinical Signs, Symptoms or Conditions	A070101	Failure to Capture		
									A070102	High Capture Threshold		
									A072202	Low Impedance		
									A1301	Failure to Read Input Signal		
							F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A070101	Failure to Capture		
									A070102	High Capture Threshold		
									A072202	Low Impedance		
									A1301	Failure to Read Input Signal		
							F2203	Imaging Required	A070101	Failure to Capture		
									A070102	High Capture Threshold		
									A072202	Low Impedance		
									A1301	Failure to Read Input Signal		
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A070101	Failure to Capture		
									A070102	High Capture Threshold		
									A072202	Low Impedance		
									A1301	Failure to Read Input Signal		
							F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A070101	Failure to Capture		
									A070102	High Capture Threshold		
									A072202	Low Impedance		
									A1301	Failure to Read Input Signal		
							F2203	Imaging Required	A070101	Failure to Capture		
									A070102	High Capture Threshold		
									A072202	Low Impedance		
									A1301	Failure to Read Input Signal		
					CT2727	Active implant devices/ancillaries	E2403	No Clinical Signs, Symptoms or Conditions	A070101	Failure to Capture		
									A070102	High Capture Threshold		
									A072202	Low Impedance		
									A1301	Failure to Read Input Signal		
							F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A070101	Failure to Capture		
									A070102	High Capture Threshold		
									A072202	Low Impedance		
									A1301	Failure to Read Input Signal		
							F2203	Imaging Required	A070101	Failure to Capture		
									A070102	High Capture Threshold		
									A072202	Low Impedance		
									A1301	Failure to Read Input Signal		
					CT2750	Implantable cardiac electrical stimulators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070101	Failure to Capture		
									A070102	High Capture Threshold		
									A072202	Low Impedance		
									A1301	Failure to Read Input Signal		
							F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A070101	Failure to Capture		
									A070102	High Capture Threshold		
									A072202	Low Impedance		
									A1301	Failure to Read Input Signal		
							F2203	Imaging Required	A070101	Failure to Capture		
									A070102	High Capture Threshold		
									A072202	Low Impedance		
									A1301	Failure to Read Input Signal		
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A070101	Failure to Capture		
									A070102	High Capture Threshold		
									A072202	Low Impedance		
									A1301	Failure to Read Input Signal		
							F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A070101	Failure to Capture		
									A070102	High Capture Threshold		
									A072202	Low Impedance		
									A1301	Failure to Read Input Signal		
							F2203	Imaging Required	A070101	Failure to Capture		
									A070102	High Capture Threshold		
									A072202	Low Impedance		
									A1301	Failure to Read Input Signal		
					CT2985	Active	E2403	No Clinical Signs, Symptoms or Conditions	A070101	Failure to Capture		
									A070102	High Capture Threshold		
									A072202	Low Impedance		
									A1301	Failure to Read Input Signal		
							F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A070101	Failure to Capture		
									A070102	High Capture Threshold		
									A072202	Low Impedance		
									A1301	Failure to Read Input Signal		
							F2203	Imaging Required	A070101	Failure to Capture		
									A070102	High Capture Threshold		
									A072202	Low Impedance		

								A1301	Failure to Read Input Signal		
					CT314	Electrically powered	E2403	No Clinical Signs, Symptoms or Conditions	A070101	Failure to Capture		
									A070102	High Capture Threshold		
									A072202	Low Impedance		
									A1301	Failure to Read Input Signal		
							F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A070101	Failure to Capture		
									A070102	High Capture Threshold		
									A072202	Low Impedance		
									A1301	Failure to Read Input Signal		
							F2203	Imaging Required	A070101	Failure to Capture		
									A070102	High Capture Threshold		
									A072202	Low Impedance		
									A1301	Failure to Read Input Signal		
12213	2023	All other reportable incidents	47263	Cardiac resynchronization therapy implantable pacemaker	CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	31/01/2023	31/01/2023
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery	31/01/2023	31/01/2023
					CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	31/01/2023	31/01/2023
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery	31/01/2023	31/01/2023
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	31/01/2023	31/01/2023
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery	31/01/2023	31/01/2023
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	31/01/2023	31/01/2023
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery	31/01/2023	31/01/2023
12214	2023	All other reportable incidents	47263	Cardiac resynchronization therapy implantable pacemaker	CT1685	Cardiac pacemakers	E011903	Syncope/Fainting	A070909	Oversensing	27/01/2023	27/01/2023
									A0712	Pacing Problem	27/01/2023	27/01/2023
									A0908	Incorrect, Inadequate or Imprecise Result or Readings	27/01/2023	27/01/2023
									A090801	Signal Artifact/Noise	27/01/2023	27/01/2023
							E060101	Asystole	A070909	Oversensing	27/01/2023	27/01/2023
									A0712	Pacing Problem	27/01/2023	27/01/2023
									A0908	Incorrect, Inadequate or Imprecise Result or Readings	27/01/2023	27/01/2023
									A090801	Signal Artifact/Noise	27/01/2023	27/01/2023
							F08	Hospitalization or Prolonged Hospitalization	A070909	Oversensing	27/01/2023	27/01/2023
									A0712	Pacing Problem	27/01/2023	27/01/2023
									A0908	Incorrect, Inadequate or Imprecise Result or Readings	27/01/2023	27/01/2023
									A090801	Signal Artifact/Noise	27/01/2023	27/01/2023
							F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A070909	Oversensing	27/01/2023	27/01/2023
									A0712	Pacing Problem	27/01/2023	27/01/2023
									A0908	Incorrect, Inadequate or Imprecise Result or Readings	27/01/2023	27/01/2023
									A090801	Signal Artifact/Noise	27/01/2023	27/01/2023
							F1905	Device Revision or Replacement	A070909	Oversensing	27/01/2023	27/01/2023
									A0712	Pacing Problem	27/01/2023	27/01/2023
									A0908	Incorrect, Inadequate or Imprecise Result or Readings	27/01/2023	27/01/2023
									A090801	Signal Artifact/Noise	27/01/2023	27/01/2023
					CT257	Implantable cardiac pacemakers	E011903	Syncope/Fainting	A070909	Oversensing	27/01/2023	27/01/2023
									A0712	Pacing Problem	27/01/2023	27/01/2023
									A0908	Incorrect, Inadequate or Imprecise Result or Readings	27/01/2023	27/01/2023
									A090801	Signal Artifact/Noise	27/01/2023	27/01/2023
							E060101	Asystole	A070909	Oversensing	27/01/2023	27/01/2023
									A0712	Pacing Problem	27/01/2023	27/01/2023
									A0908	Incorrect, Inadequate or Imprecise Result or Readings	27/01/2023	27/01/2023
									A090801	Signal Artifact/Noise	27/01/2023	27/01/2023
							F08	Hospitalization or Prolonged Hospitalization	A070909	Oversensing	27/01/2023	27/01/2023
									A0712	Pacing Problem	27/01/2023	27/01/2023
									A0908	Incorrect, Inadequate or Imprecise Result or Readings	27/01/2023	27/01/2023
									A090801	Signal Artifact/Noise	27/01/2023	27/01/2023
							F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A070909	Oversensing	27/01/2023	27/01/2023
									A0712	Pacing Problem	27/01/2023	27/01/2023
									A0908	Incorrect, Inadequate or Imprecise Result or Readings	27/01/2023	27/01/2023
									A090801	Signal Artifact/Noise	27/01/2023	27/01/2023
							F1905	Device Revision or Replacement	A070909	Oversensing	27/01/2023	27/01/2023
									A0712	Pacing Problem	27/01/2023	27/01/2023
									A0908	Incorrect, Inadequate or Imprecise Result or Readings	27/01/2023	27/01/2023
									A090801	Signal Artifact/Noise	27/01/2023	27/01/2023
					CT2866	Implantable cardiac electrical stimulators	E011903	Syncope/Fainting	A070909	Oversensing	27/01/2023	27/01/2023
									A0712	Pacing Problem	27/01/2023	27/01/2023
									A0908	Incorrect, Inadequate or Imprecise Result or Readings	27/01/2023	27/01/2023
									A090801	Signal Artifact/Noise	27/01/2023	27/01/2023
							E060101	Asystole	A070909	Oversensing	27/01/2023	27/01/2023
									A0712	Pacing Problem	27/01/2023	27/01/2023
									A0908	Incorrect, Inadequate or Imprecise Result or Readings	27/01/2023	27/01/2023
									A090801	Signal Artifact/Noise	27/01/2023	27/01/2023
							F08	Hospitalization or Prolonged Hospitalization	A070909	Oversensing	27/01/2023	27/01/2023
									A0712	Pacing Problem	27/01/2023	27/01/2023
									A0908	Incorrect, Inadequate or Imprecise Result or Readings	27/01/2023	27/01/2023
									A090801	Signal Artifact/Noise	27/01/2023	27/01/2023
							F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A070909	Oversensing	27/01/2023	27/01/2023
									A0712	Pacing Problem	27/01/2023	27/01/2023
									A0908	Incorrect, Inadequate or Imprecise Result or Readings	27/01/2023	27/01/2023
									A090801	Signal Artifact/Noise	27/01/2023	27/01/2023

							F1905	Device Revision or Replacement	A070909	Oversensing	27/01/2023	27/01/2023
									A0712	Pacing Problem	27/01/2023	27/01/2023
									A0908	Incorrect, Inadequate or Imprecise Result or Readings	27/01/2023	27/01/2023
									A090801	Signal Artifact/Noise	27/01/2023	27/01/2023
12215	2023	All other reportable incidents	47265	Dual-chamber implantable pacemaker, rate-responsive	CT141	Electrodes/Leads and associated devices	E011902	Presyncope	A0712	Pacing Problem	01/02/2023	01/02/2023
							E011903	Syncope/Fainting	A0712	Pacing Problem	01/02/2023	01/02/2023
							F08	Hospitalization or Prolonged Hospitalization	A0712	Pacing Problem	01/02/2023	01/02/2023
							F1203	Life Threatening Illness or Injury	A0712	Pacing Problem	01/02/2023	01/02/2023
							F1903	Device Explantation	A0712	Pacing Problem	01/02/2023	01/02/2023
					CT145	Cardiovascular devices	E011902	Presyncope	A0712	Pacing Problem	01/02/2023	01/02/2023
							E011903	Syncope/Fainting	A0712	Pacing Problem	01/02/2023	01/02/2023
							F08	Hospitalization or Prolonged Hospitalization	A0712	Pacing Problem	01/02/2023	01/02/2023
							F1203	Life Threatening Illness or Injury	A0712	Pacing Problem	01/02/2023	01/02/2023
							F1903	Device Explantation	A0712	Pacing Problem	01/02/2023	01/02/2023
					CT1684	Cardiac pacemakers and associated devices	E011902	Presyncope	A0712	Pacing Problem	01/02/2023	01/02/2023
							E011903	Syncope/Fainting	A0712	Pacing Problem	01/02/2023	01/02/2023
							F08	Hospitalization or Prolonged Hospitalization	A0712	Pacing Problem	01/02/2023	01/02/2023
							F1203	Life Threatening Illness or Injury	A0712	Pacing Problem	01/02/2023	01/02/2023
							F1903	Device Explantation	A0712	Pacing Problem	01/02/2023	01/02/2023
					CT1685	Cardiac pacemakers	E011902	Presyncope	A0712	Pacing Problem	01/02/2023	01/02/2023
							E011903	Syncope/Fainting	A0712	Pacing Problem	01/02/2023	01/02/2023
							F08	Hospitalization or Prolonged Hospitalization	A0712	Pacing Problem	01/02/2023	01/02/2023
							F1203	Life Threatening Illness or Injury	A0712	Pacing Problem	01/02/2023	01/02/2023
							F1903	Device Explantation	A0712	Pacing Problem	01/02/2023	01/02/2023
					CT233	Surgical	E011902	Presyncope	A0712	Pacing Problem	01/02/2023	01/02/2023
							E011903	Syncope/Fainting	A0712	Pacing Problem	01/02/2023	01/02/2023
							F08	Hospitalization or Prolonged Hospitalization	A0712	Pacing Problem	01/02/2023	01/02/2023
							F1203	Life Threatening Illness or Injury	A0712	Pacing Problem	01/02/2023	01/02/2023
							F1903	Device Explantation	A0712	Pacing Problem	01/02/2023	01/02/2023
					CT2406	Implantable	E011902	Presyncope	A0712	Pacing Problem	01/02/2023	01/02/2023
							E011903	Syncope/Fainting	A0712	Pacing Problem	01/02/2023	01/02/2023
							F08	Hospitalization or Prolonged Hospitalization	A0712	Pacing Problem	01/02/2023	01/02/2023
							F1203	Life Threatening Illness or Injury	A0712	Pacing Problem	01/02/2023	01/02/2023
							F1903	Device Explantation	A0712	Pacing Problem	01/02/2023	01/02/2023
					CT2407	Active implantable	E011902	Presyncope	A0712	Pacing Problem	01/02/2023	01/02/2023
							E011903	Syncope/Fainting	A0712	Pacing Problem	01/02/2023	01/02/2023
							F08	Hospitalization or Prolonged Hospitalization	A0712	Pacing Problem	01/02/2023	01/02/2023
							F1203	Life Threatening Illness or Injury	A0712	Pacing Problem	01/02/2023	01/02/2023
							F1903	Device Explantation	A0712	Pacing Problem	01/02/2023	01/02/2023
					CT257	Implantable cardiac pacemakers	E011902	Presyncope	A0712	Pacing Problem	01/02/2023	01/02/2023
							E011903	Syncope/Fainting	A0712	Pacing Problem	01/02/2023	01/02/2023
							F08	Hospitalization or Prolonged Hospitalization	A0712	Pacing Problem	01/02/2023	01/02/2023
							F1203	Life Threatening Illness or Injury	A0712	Pacing Problem	01/02/2023	01/02/2023
							F1903	Device Explantation	A0712	Pacing Problem	01/02/2023	01/02/2023
					CT2727	Active implant devices/ancillaries	E011902	Presyncope	A0712	Pacing Problem	01/02/2023	01/02/2023
							E011903	Syncope/Fainting	A0712	Pacing Problem	01/02/2023	01/02/2023
							F08	Hospitalization or Prolonged Hospitalization	A0712	Pacing Problem	01/02/2023	01/02/2023
							F1203	Life Threatening Illness or Injury	A0712	Pacing Problem	01/02/2023	01/02/2023
							F1903	Device Explantation	A0712	Pacing Problem	01/02/2023	01/02/2023
					CT2750	Implantable cardiac electrical stimulators and associated devices	E011902	Presyncope	A0712	Pacing Problem	01/02/2023	01/02/2023
							E011903	Syncope/Fainting	A0712	Pacing Problem	01/02/2023	01/02/2023
							F08	Hospitalization or Prolonged Hospitalization	A0712	Pacing Problem	01/02/2023	01/02/2023
							F1203	Life Threatening Illness or Injury	A0712	Pacing Problem	01/02/2023	01/02/2023
							F1903	Device Explantation	A0712	Pacing Problem	01/02/2023	01/02/2023
					CT2866	Implantable cardiac electrical stimulators	E011902	Presyncope	A0712	Pacing Problem	01/02/2023	01/02/2023
							E011903	Syncope/Fainting	A0712	Pacing Problem	01/02/2023	01/02/2023
							F08	Hospitalization or Prolonged Hospitalization	A0712	Pacing Problem	01/02/2023	01/02/2023
							F1203	Life Threatening Illness or Injury	A0712	Pacing Problem	01/02/2023	01/02/2023
							F1903	Device Explantation	A0712	Pacing Problem	01/02/2023	01/02/2023
					CT2985	Active	E011902	Presyncope	A0712	Pacing Problem	01/02/2023	01/02/2023
							E011903	Syncope/Fainting	A0712	Pacing Problem	01/02/2023	01/02/2023
							F08	Hospitalization or Prolonged Hospitalization	A0712	Pacing Problem	01/02/2023	01/02/2023
							F1203	Life Threatening Illness or Injury	A0712	Pacing Problem	01/02/2023	01/02/2023
							F1903	Device Explantation	A0712	Pacing Problem	01/02/2023	01/02/2023
					CT3001	Implantable metal	E011902	Presyncope	A0712	Pacing Problem	01/02/2023	01/02/2023
							E011903	Syncope/Fainting	A0712	Pacing Problem	01/02/2023	01/02/2023
							F08	Hospitalization or Prolonged Hospitalization	A0712	Pacing Problem	01/02/2023	01/02/2023
							F1203	Life Threatening Illness or Injury	A0712	Pacing Problem	01/02/2023	01/02/2023
							F1903	Device Explantation	A0712	Pacing Problem	01/02/2023	01/02/2023
					CT3002	Implantable synthetic polymer	E011902	Presyncope	A0712	Pacing Problem	01/02/2023	01/02/2023
							E011903	Syncope/Fainting	A0712	Pacing Problem	01/02/2023	01/02/2023
							F08	Hospitalization or Prolonged Hospitalization	A0712	Pacing Problem	01/02/2023	01/02/2023
							F1203	Life Threatening Illness or Injury	A0712	Pacing Problem	01/02/2023	01/02/2023
							F1903	Device Explantation	A0712	Pacing Problem	01/02/2023	01/02/2023
					CT3006	Implantable inorganic material	E011902	Presyncope	A0712	Pacing Problem	01/02/2023	01/02/2023
							E011903	Syncope/Fainting	A0712	Pacing Problem	01/02/2023	01/02/2023
							F08	Hospitalization or Prolonged Hospitalization	A0712	Pacing Problem	01/02/2023	01/02/2023
							F1203	Life Threatening Illness or Injury	A0712	Pacing Problem	01/02/2023	01/02/2023
							F1903	Device Explantation	A0712	Pacing Problem	01/02/2023	01/02/2023
					CT314	Electrically powered	E011902	Presyncope	A0712	Pacing Problem	01/02/2023	01/02/2023
							E011903	Syncope/Fainting	A0712	Pacing Problem	01/02/2023	01/02/2023
							F08	Hospitalization or Prolonged Hospitalization	A0712	Pacing Problem	01/02/2023	01/02/2023
							F1203	Life Threatening Illness or Injury	A0712	Pacing Problem	01/02/2023	01/02/2023
							F1903	Device Explantation	A0712	Pacing Problem	01/02/2023	01/02/2023
					CT321	Long-term surgical invasive	E011902	Presyncope	A0712	Pacing Problem	01/02/2023	01/02/2023
							E011903	Syncope/Fainting	A0712	Pacing Problem	01/02/2023	01/02/2023
							F08	Hospitalization or Prolonged Hospitalization	A0712	Pacing Problem	01/02/2023	01/02/2023

							F1203	Life Threatening Illness or Injury	A0712	Pacing Problem	01/02/2023	01/02/2023
							F1903	Device Explantation	A0712	Pacing Problem	01/02/2023	01/02/2023
					CT449	Generators and associated devices	E011902	Presyncope	A0712	Pacing Problem	01/02/2023	01/02/2023
							E011903	Syncope/Fainting	A0712	Pacing Problem	01/02/2023	01/02/2023
							F08	Hospitalization or Prolonged Hospitalization	A0712	Pacing Problem	01/02/2023	01/02/2023
							F1203	Life Threatening Illness or Injury	A0712	Pacing Problem	01/02/2023	01/02/2023
							F1903	Device Explantation	A0712	Pacing Problem	01/02/2023	01/02/2023
					CT981	Single use	E011902	Presyncope	A0712	Pacing Problem	01/02/2023	01/02/2023
							E011903	Syncope/Fainting	A0712	Pacing Problem	01/02/2023	01/02/2023
							F08	Hospitalization or Prolonged Hospitalization	A0712	Pacing Problem	01/02/2023	01/02/2023
							F1203	Life Threatening Illness or Injury	A0712	Pacing Problem	01/02/2023	01/02/2023
							F1903	Device Explantation	A0712	Pacing Problem	01/02/2023	01/02/2023
					CT983	Surgical invasive	E011902	Presyncope	A0712	Pacing Problem	01/02/2023	01/02/2023
							E011903	Syncope/Fainting	A0712	Pacing Problem	01/02/2023	01/02/2023
							F08	Hospitalization or Prolonged Hospitalization	A0712	Pacing Problem	01/02/2023	01/02/2023
							F1203	Life Threatening Illness or Injury	A0712	Pacing Problem	01/02/2023	01/02/2023
							F1903	Device Explantation	A0712	Pacing Problem	01/02/2023	01/02/2023
12216	2023	All other reportable incidents	47265	Dual-chamber implantable pacemaker, rate-responsive	CT141	Electrodes/Leads and associated devices	E2401	Insufficient Information	A0501	Detachment of Device or Device Component	13/10/2021	
									A070101	Failure to Capture	13/10/2021	
							F24	Insufficient Health Impact Information	A0501	Detachment of Device or Device Component	13/10/2021	
									A070101	Failure to Capture	13/10/2021	
					CT145	Cardiovascular devices	E2401	Insufficient Information	A0501	Detachment of Device or Device Component	13/10/2021	
									A070101	Failure to Capture	13/10/2021	
							F24	Insufficient Health Impact Information	A0501	Detachment of Device or Device Component	13/10/2021	
									A070101	Failure to Capture	13/10/2021	
					CT1684	Cardiac pacemakers and associated devices	E2401	Insufficient Information	A0501	Detachment of Device or Device Component	13/10/2021	
									A070101	Failure to Capture	13/10/2021	
							F24	Insufficient Health Impact Information	A0501	Detachment of Device or Device Component	13/10/2021	
									A070101	Failure to Capture	13/10/2021	
					CT1685	Cardiac pacemakers	E2401	Insufficient Information	A0501	Detachment of Device or Device Component	13/10/2021	
									A070101	Failure to Capture	13/10/2021	
							F24	Insufficient Health Impact Information	A0501	Detachment of Device or Device Component	13/10/2021	
									A070101	Failure to Capture	13/10/2021	
					CT233	Surgical	E2401	Insufficient Information	A0501	Detachment of Device or Device Component	13/10/2021	
									A070101	Failure to Capture	13/10/2021	
							F24	Insufficient Health Impact Information	A0501	Detachment of Device or Device Component	13/10/2021	
									A070101	Failure to Capture	13/10/2021	
					CT2406	Implantable	E2401	Insufficient Information	A0501	Detachment of Device or Device Component	13/10/2021	
									A070101	Failure to Capture	13/10/2021	
							F24	Insufficient Health Impact Information	A0501	Detachment of Device or Device Component	13/10/2021	
									A070101	Failure to Capture	13/10/2021	
					CT2407	Active implantable	E2401	Insufficient Information	A0501	Detachment of Device or Device Component	13/10/2021	
									A070101	Failure to Capture	13/10/2021	
							F24	Insufficient Health Impact Information	A0501	Detachment of Device or Device Component	13/10/2021	
									A070101	Failure to Capture	13/10/2021	
					CT257	Implantable cardiac pacemakers	E2401	Insufficient Information	A0501	Detachment of Device or Device Component	13/10/2021	
									A070101	Failure to Capture	13/10/2021	
							F24	Insufficient Health Impact Information	A0501	Detachment of Device or Device Component	13/10/2021	
									A070101	Failure to Capture	13/10/2021	
					CT2727	Active implant devices/ancillaries	E2401	Insufficient Information	A0501	Detachment of Device or Device Component	13/10/2021	
									A070101	Failure to Capture	13/10/2021	
							F24	Insufficient Health Impact Information	A0501	Detachment of Device or Device Component	13/10/2021	
									A070101	Failure to Capture	13/10/2021	
					CT2750	Implantable cardiac electrical stimulators and associated devices	E2401	Insufficient Information	A0501	Detachment of Device or Device Component	13/10/2021	
									A070101	Failure to Capture	13/10/2021	
							F24	Insufficient Health Impact Information	A0501	Detachment of Device or Device Component	13/10/2021	
									A070101	Failure to Capture	13/10/2021	
					CT2866	Implantable cardiac electrical stimulators	E2401	Insufficient Information	A0501	Detachment of Device or Device Component	13/10/2021	
									A070101	Failure to Capture	13/10/2021	
							F24	Insufficient Health Impact Information	A0501	Detachment of Device or Device Component	13/10/2021	
									A070101	Failure to Capture	13/10/2021	
					CT2985	Active	E2401	Insufficient Information	A0501	Detachment of Device or Device Component	13/10/2021	
									A070101	Failure to Capture	13/10/2021	
							F24	Insufficient Health Impact Information	A0501	Detachment of Device or Device Component	13/10/2021	
									A070101	Failure to Capture	13/10/2021	
					CT3001	Implantable metal	E2401	Insufficient Information	A0501	Detachment of Device or Device Component	13/10/2021	
									A070101	Failure to Capture	13/10/2021	
							F24	Insufficient Health Impact Information	A0501	Detachment of Device or Device Component	13/10/2021	
									A070101	Failure to Capture	13/10/2021	
					CT3002	Implantable synthetic polymer	E2401	Insufficient Information	A0501	Detachment of Device or Device Component	13/10/2021	
									A070101	Failure to Capture	13/10/2021	
							F24	Insufficient Health Impact Information	A0501	Detachment of Device or Device Component	13/10/2021	
									A070101	Failure to Capture	13/10/2021	
					CT3006	Implantable inorganic material	E2401	Insufficient Information	A0501	Detachment of Device or Device Component	13/10/2021	
									A070101	Failure to Capture	13/10/2021	
							F24	Insufficient Health Impact Information	A0501	Detachment of Device or Device Component	13/10/2021	
									A070101	Failure to Capture	13/10/2021	
					CT314	Electrically powered	E2401	Insufficient Information	A0501	Detachment of Device or Device Component	13/10/2021	
									A070101	Failure to Capture	13/10/2021	
							F24	Insufficient Health Impact Information	A0501	Detachment of Device or Device Component	13/10/2021	
									A070101	Failure to Capture	13/10/2021	
					CT321	Long-term surgical invasive	E2401	Insufficient Information	A0501	Detachment of Device or Device Component	13/10/2021	
									A070101	Failure to Capture	13/10/2021	
							F24	Insufficient Health Impact Information	A0501	Detachment of Device or Device Component	13/10/2021	
									A070101	Failure to Capture	13/10/2021	
					CT449	Generators and associated devices	E2401	Insufficient Information	A0501	Detachment of Device or Device Component	13/10/2021	
									A070101	Failure to Capture	13/10/2021	
							F24	Insufficient Health Impact Information	A0501	Detachment of Device or Device Component	13/10/2021	

									A070101	Failure to Capture	13/10/2021	
					CT981	Single use	E2401	Insufficient Information	A0501	Detachment of Device or Device Component	13/10/2021	
									A070101	Failure to Capture	13/10/2021	
							F24	Insufficient Health Impact Information	A0501	Detachment of Device or Device Component	13/10/2021	
									A070101	Failure to Capture	13/10/2021	
					CT983	Surgical invasive	E2401	Insufficient Information	A0501	Detachment of Device or Device Component	13/10/2021	
									A070101	Failure to Capture	13/10/2021	
							F24	Insufficient Health Impact Information	A0501	Detachment of Device or Device Component	13/10/2021	
									A070101	Failure to Capture	13/10/2021	
12217	2023	All other reportable incidents	57941	Cardiovascular device introducer, non-steerable	CT1684	Cardiac pacemakers and associated devices	E1002	Abdominal Pain	A04	Material Integrity Problem		
									A0414	Material Split, Cut or Torn		
									A150204	Failure to Advance		
							F1908	Prolonged Surgery	A04	Material Integrity Problem		
									A0414	Material Split, Cut or Torn		
									A150204	Failure to Advance		
							F1909	Surgical Procedure Delayed	A04	Material Integrity Problem		
									A0414	Material Split, Cut or Torn		
									A150204	Failure to Advance		
12218	2023	All other reportable incidents	37265	Dual-chamber implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A070909	Oversensing		
									A090801	Signal Artifact/Noise		
							F26	No Health Consequences or Impact	A070909	Oversensing		
									A090801	Signal Artifact/Noise		
12219	2023		47265	Dual-chamber implantable pacemaker, rate-responsive	CT141	Electrodes/Leads and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem		
							F1901	Additional Surgery	A24	No Apparent Device Problem		
					CT145	Cardiovascular devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem		
							F1901	Additional Surgery	A24	No Apparent Device Problem		
					CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem		
							F1901	Additional Surgery	A24	No Apparent Device Problem		
					CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem		
							F1901	Additional Surgery	A24	No Apparent Device Problem		
					CT233	Surgical	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem		
							F1901	Additional Surgery	A24	No Apparent Device Problem		
					CT2406	Implantable	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem		
							F1901	Additional Surgery	A24	No Apparent Device Problem		
					CT2407	Active implantable	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem		
							F1901	Additional Surgery	A24	No Apparent Device Problem		
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem		
							F1901	Additional Surgery	A24	No Apparent Device Problem		
					CT2727	Active implant devices/ancillaries	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem		
							F1901	Additional Surgery	A24	No Apparent Device Problem		
					CT2750	Implantable cardiac electrical stimulators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem		
							F1901	Additional Surgery	A24	No Apparent Device Problem		
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem		
							F1901	Additional Surgery	A24	No Apparent Device Problem		
					CT2985	Active	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem		
							F1901	Additional Surgery	A24	No Apparent Device Problem		
					CT3001	Implantable metal	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem		
							F1901	Additional Surgery	A24	No Apparent Device Problem		
					CT3002	Implantable synthetic polymer	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem		
							F1901	Additional Surgery	A24	No Apparent Device Problem		
					CT3006	Implantable inorganic material	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem		
							F1901	Additional Surgery	A24	No Apparent Device Problem		
					CT314	Electrically powered	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem		
							F1901	Additional Surgery	A24	No Apparent Device Problem		
					CT321	Long-term surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem		
							F1901	Additional Surgery	A24	No Apparent Device Problem		
					CT449	Generators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem		
							F1901	Additional Surgery	A24	No Apparent Device Problem		
					CT981	Single use	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem		
							F1901	Additional Surgery	A24	No Apparent Device Problem		
					CT983	Surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem		
							F1901	Additional Surgery	A24	No Apparent Device Problem		
12220	2023	All other reportable incidents	36241	Pacemaker/defibrillator lead	CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A072201	High Impedance		
							F26	No Health Consequences or Impact	A072201	High Impedance		
12221	2023	All other reportable incidents	47265	Dual-chamber implantable pacemaker, rate-responsive	CT141	Electrodes/Leads and associated devices	E2340	Wound Dehiscence	A0104	Migration or Expulsion of Device	08/11/2022	08/11/2022
									A24	No Apparent Device Problem	08/11/2022	08/11/2022
							F1901	Additional Surgery	A0104	Migration or Expulsion of Device	08/11/2022	08/11/2022
									A24	No Apparent Device Problem	08/11/2022	08/11/2022
					CT145	Cardiovascular devices	E2340	Wound Dehiscence	A0104	Migration or Expulsion of Device	08/11/2022	08/11/2022
									A24	No Apparent Device Problem	08/11/2022	08/11/2022
							F1901	Additional Surgery	A0104	Migration or Expulsion of Device	08/11/2022	08/11/2022
									A24	No Apparent Device Problem	08/11/2022	08/11/2022
					CT1684	Cardiac pacemakers and associated devices	E2340	Wound Dehiscence	A0104	Migration or Expulsion of Device	08/11/2022	08/11/2022
									A24	No Apparent Device Problem	08/11/2022	08/11/2022
							F1901	Additional Surgery	A0104	Migration or Expulsion of Device	08/11/2022	08/11/2022
									A24	No Apparent Device Problem	08/11/2022	08/11/2022
					CT1685	Cardiac pacemakers	E2340	Wound Dehiscence	A0104	Migration or Expulsion of Device	08/11/2022	08/11/2022
									A24	No Apparent Device Problem	08/11/2022	08/11/2022
							F1901	Additional Surgery	A0104	Migration or Expulsion of Device	08/11/2022	08/11/2022
									A24	No Apparent Device Problem	08/11/2022	08/11/2022
					CT233	Surgical	E2340	Wound Dehiscence	A0104	Migration or Expulsion of Device	08/11/2022	08/11/2022
									A24	No Apparent Device Problem	08/11/2022	08/11/2022
							F1901	Additional Surgery	A0104	Migration or Expulsion of Device	08/11/2022	08/11/2022
									A24	No Apparent Device Problem	08/11/2022	08/11/2022
					CT2406	Implantable	E2340	Wound Dehiscence	A0104	Migration or Expulsion of Device	08/11/2022	08/11/2022
									A24	No Apparent Device Problem	08/11/2022	08/11/2022
							F1901	Additional Surgery	A0104	Migration or Expulsion of Device	08/11/2022	08/11/2022
									A24	No Apparent Device Problem	08/11/2022	08/11/2022

					CT2407	Active implantable	E2340	Wound Dehiscence	A0104	Migration or Expulsion of Device	08/11/2022	08/11/2022
									A24	No Apparent Device Problem	08/11/2022	08/11/2022
							F1901	Additional Surgery	A0104	Migration or Expulsion of Device	08/11/2022	08/11/2022
									A24	No Apparent Device Problem	08/11/2022	08/11/2022
					CT257	Implantable cardiac pacemakers	E2340	Wound Dehiscence	A0104	Migration or Expulsion of Device	08/11/2022	08/11/2022
									A24	No Apparent Device Problem	08/11/2022	08/11/2022
							F1901	Additional Surgery	A0104	Migration or Expulsion of Device	08/11/2022	08/11/2022
									A24	No Apparent Device Problem	08/11/2022	08/11/2022
					CT2727	Active implant devices/ancillaries	E2340	Wound Dehiscence	A0104	Migration or Expulsion of Device	08/11/2022	08/11/2022
									A24	No Apparent Device Problem	08/11/2022	08/11/2022
							F1901	Additional Surgery	A0104	Migration or Expulsion of Device	08/11/2022	08/11/2022
									A24	No Apparent Device Problem	08/11/2022	08/11/2022
					CT2750	Implantable cardiac electrical stimulators and associated devices	E2340	Wound Dehiscence	A0104	Migration or Expulsion of Device	08/11/2022	08/11/2022
									A24	No Apparent Device Problem	08/11/2022	08/11/2022
							F1901	Additional Surgery	A0104	Migration or Expulsion of Device	08/11/2022	08/11/2022
									A24	No Apparent Device Problem	08/11/2022	08/11/2022
					CT2866	Implantable cardiac electrical stimulators	E2340	Wound Dehiscence	A0104	Migration or Expulsion of Device	08/11/2022	08/11/2022
									A24	No Apparent Device Problem	08/11/2022	08/11/2022
							F1901	Additional Surgery	A0104	Migration or Expulsion of Device	08/11/2022	08/11/2022
									A24	No Apparent Device Problem	08/11/2022	08/11/2022
					CT2985	Active	E2340	Wound Dehiscence	A0104	Migration or Expulsion of Device	08/11/2022	08/11/2022
									A24	No Apparent Device Problem	08/11/2022	08/11/2022
							F1901	Additional Surgery	A0104	Migration or Expulsion of Device	08/11/2022	08/11/2022
									A24	No Apparent Device Problem	08/11/2022	08/11/2022
					CT3001	Implantable metal	E2340	Wound Dehiscence	A0104	Migration or Expulsion of Device	08/11/2022	08/11/2022
									A24	No Apparent Device Problem	08/11/2022	08/11/2022
							F1901	Additional Surgery	A0104	Migration or Expulsion of Device	08/11/2022	08/11/2022
									A24	No Apparent Device Problem	08/11/2022	08/11/2022
					CT3002	Implantable synthetic polymer	E2340	Wound Dehiscence	A0104	Migration or Expulsion of Device	08/11/2022	08/11/2022
									A24	No Apparent Device Problem	08/11/2022	08/11/2022
							F1901	Additional Surgery	A0104	Migration or Expulsion of Device	08/11/2022	08/11/2022
									A24	No Apparent Device Problem	08/11/2022	08/11/2022
					CT3006	Implantable inorganic material	E2340	Wound Dehiscence	A0104	Migration or Expulsion of Device	08/11/2022	08/11/2022
									A24	No Apparent Device Problem	08/11/2022	08/11/2022
							F1901	Additional Surgery	A0104	Migration or Expulsion of Device	08/11/2022	08/11/2022
									A24	No Apparent Device Problem	08/11/2022	08/11/2022
					CT314	Electrically powered	E2340	Wound Dehiscence	A0104	Migration or Expulsion of Device	08/11/2022	08/11/2022
									A24	No Apparent Device Problem	08/11/2022	08/11/2022
							F1901	Additional Surgery	A0104	Migration or Expulsion of Device	08/11/2022	08/11/2022
									A24	No Apparent Device Problem	08/11/2022	08/11/2022
					CT321	Long-term surgical invasive	E2340	Wound Dehiscence	A0104	Migration or Expulsion of Device	08/11/2022	08/11/2022
									A24	No Apparent Device Problem	08/11/2022	08/11/2022
							F1901	Additional Surgery	A0104	Migration or Expulsion of Device	08/11/2022	08/11/2022
									A24	No Apparent Device Problem	08/11/2022	08/11/2022
					CT449	Generators and associated devices	E2340	Wound Dehiscence	A0104	Migration or Expulsion of Device	08/11/2022	08/11/2022
									A24	No Apparent Device Problem	08/11/2022	08/11/2022
							F1901	Additional Surgery	A0104	Migration or Expulsion of Device	08/11/2022	08/11/2022
									A24	No Apparent Device Problem	08/11/2022	08/11/2022
					CT981	Single use	E2340	Wound Dehiscence	A0104	Migration or Expulsion of Device	08/11/2022	08/11/2022
									A24	No Apparent Device Problem	08/11/2022	08/11/2022
							F1901	Additional Surgery	A0104	Migration or Expulsion of Device	08/11/2022	08/11/2022
									A24	No Apparent Device Problem	08/11/2022	08/11/2022
					CT983	Surgical invasive	E2340	Wound Dehiscence	A0104	Migration or Expulsion of Device	08/11/2022	08/11/2022
									A24	No Apparent Device Problem	08/11/2022	08/11/2022
							F1901	Additional Surgery	A0104	Migration or Expulsion of Device	08/11/2022	08/11/2022
									A24	No Apparent Device Problem	08/11/2022	08/11/2022
12222	2023		47265	Dual-chamber implantable pacemaker, rate-responsive	CT141	Electrodes/Leads and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	30/01/2023	30/01/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	30/01/2023	30/01/2023
					CT145	Cardiovascular devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	30/01/2023	30/01/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	30/01/2023	30/01/2023
					CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	30/01/2023	30/01/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	30/01/2023	30/01/2023
					CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	30/01/2023	30/01/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	30/01/2023	30/01/2023
					CT233	Surgical	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	30/01/2023	30/01/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	30/01/2023	30/01/2023
					CT2406	Implantable	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	30/01/2023	30/01/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	30/01/2023	30/01/2023
					CT2407	Active implantable	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	30/01/2023	30/01/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	30/01/2023	30/01/2023
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	30/01/2023	30/01/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	30/01/2023	30/01/2023
					CT2727	Active implant devices/ancillaries	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	30/01/2023	30/01/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	30/01/2023	30/01/2023
					CT2750	Implantable cardiac electrical stimulators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	30/01/2023	30/01/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	30/01/2023	30/01/2023
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	30/01/2023	30/01/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	30/01/2023	30/01/2023
					CT2985	Active	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	30/01/2023	30/01/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	30/01/2023	30/01/2023
					CT3001	Implantable metal	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	30/01/2023	30/01/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	30/01/2023	30/01/2023
					CT3002	Implantable synthetic polymer	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	30/01/2023	30/01/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	30/01/2023	30/01/2023
					CT3006	Implantable inorganic material	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	30/01/2023	30/01/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	30/01/2023	30/01/2023
					CT314	Electrically powered	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	30/01/2023	30/01/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	30/01/2023	30/01/2023

					CT321	Long-term surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	30/01/2023	30/01/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	30/01/2023	30/01/2023
					CT449	Generators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	30/01/2023	30/01/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	30/01/2023	30/01/2023
					CT981	Single use	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	30/01/2023	30/01/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	30/01/2023	30/01/2023
					CT983	Surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	30/01/2023	30/01/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	30/01/2023	30/01/2023
12223	2023		47265	Dual-chamber implantable pacemaker, rate-responsive	CT141	Electrodes/Leads and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	24/01/2023	24/01/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	24/01/2023	24/01/2023
					CT145	Cardiovascular devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	24/01/2023	24/01/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	24/01/2023	24/01/2023
					CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	24/01/2023	24/01/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	24/01/2023	24/01/2023
					CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	24/01/2023	24/01/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	24/01/2023	24/01/2023
					CT233	Surgical	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	24/01/2023	24/01/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	24/01/2023	24/01/2023
					CT2406	Implantable	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	24/01/2023	24/01/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	24/01/2023	24/01/2023
					CT2407	Active implantable	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	24/01/2023	24/01/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	24/01/2023	24/01/2023
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	24/01/2023	24/01/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	24/01/2023	24/01/2023
					CT2727	Active implant devices/ancillaries	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	24/01/2023	24/01/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	24/01/2023	24/01/2023
					CT2750	Implantable cardiac electrical stimulators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	24/01/2023	24/01/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	24/01/2023	24/01/2023
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	24/01/2023	24/01/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	24/01/2023	24/01/2023
					CT2985	Active	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	24/01/2023	24/01/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	24/01/2023	24/01/2023
					CT3001	Implantable metal	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	24/01/2023	24/01/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	24/01/2023	24/01/2023
					CT3002	Implantable synthetic polymer	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	24/01/2023	24/01/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	24/01/2023	24/01/2023
					CT3006	Implantable inorganic material	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	24/01/2023	24/01/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	24/01/2023	24/01/2023
					CT314	Electrically powered	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	24/01/2023	24/01/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	24/01/2023	24/01/2023
					CT321	Long-term surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	24/01/2023	24/01/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	24/01/2023	24/01/2023
					CT449	Generators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	24/01/2023	24/01/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	24/01/2023	24/01/2023
					CT981	Single use	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	24/01/2023	24/01/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	24/01/2023	24/01/2023
					CT983	Surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	24/01/2023	24/01/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	24/01/2023	24/01/2023
12224	2023		47265	Dual-chamber implantable pacemaker, rate-responsive	CT141	Electrodes/Leads and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	31/01/2023	31/01/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	31/01/2023	31/01/2023
					CT145	Cardiovascular devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	31/01/2023	31/01/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	31/01/2023	31/01/2023
					CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	31/01/2023	31/01/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	31/01/2023	31/01/2023
					CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	31/01/2023	31/01/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	31/01/2023	31/01/2023
					CT233	Surgical	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	31/01/2023	31/01/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	31/01/2023	31/01/2023
					CT2406	Implantable	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	31/01/2023	31/01/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	31/01/2023	31/01/2023
					CT2407	Active implantable	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	31/01/2023	31/01/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	31/01/2023	31/01/2023
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	31/01/2023	31/01/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	31/01/2023	31/01/2023
					CT2727	Active implant devices/ancillaries	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	31/01/2023	31/01/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	31/01/2023	31/01/2023
					CT2750	Implantable cardiac electrical stimulators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	31/01/2023	31/01/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	31/01/2023	31/01/2023
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	31/01/2023	31/01/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	31/01/2023	31/01/2023
					CT2985	Active	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	31/01/2023	31/01/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	31/01/2023	31/01/2023
					CT3001	Implantable metal	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	31/01/2023	31/01/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	31/01/2023	31/01/2023
					CT3002	Implantable synthetic polymer	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	31/01/2023	31/01/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	31/01/2023	31/01/2023
					CT3006	Implantable inorganic material	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	31/01/2023	31/01/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	31/01/2023	31/01/2023
					CT314	Electrically powered	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	31/01/2023	31/01/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	31/01/2023	31/01/2023
					CT321	Long-term surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	31/01/2023	31/01/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	31/01/2023	31/01/2023
					CT449	Generators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	31/01/2023	31/01/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	31/01/2023	31/01/2023
					CT981	Single use	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	31/01/2023	31/01/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	31/01/2023	31/01/2023
					CT983	Surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	31/01/2023	31/01/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	31/01/2023	31/01/2023

12225	2023	All other reportable incidents	47265	Dual-chamber implantable pacemaker, rate-responsive	CT141	Electrodes/Leads and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	08/01/2023	08/01/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	08/01/2023	08/01/2023
					CT145	Cardiovascular devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	08/01/2023	08/01/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	08/01/2023	08/01/2023
					CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	08/01/2023	08/01/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	08/01/2023	08/01/2023
					CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	08/01/2023	08/01/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	08/01/2023	08/01/2023
					CT233	Surgical	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	08/01/2023	08/01/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	08/01/2023	08/01/2023
					CT2406	Implantable	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	08/01/2023	08/01/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	08/01/2023	08/01/2023
					CT2407	Active implantable	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	08/01/2023	08/01/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	08/01/2023	08/01/2023
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	08/01/2023	08/01/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	08/01/2023	08/01/2023
					CT2727	Active implant devices/ancillaries	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	08/01/2023	08/01/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	08/01/2023	08/01/2023
					CT2750	Implantable cardiac electrical stimulators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	08/01/2023	08/01/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	08/01/2023	08/01/2023
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	08/01/2023	08/01/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	08/01/2023	08/01/2023
					CT2985	Active	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	08/01/2023	08/01/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	08/01/2023	08/01/2023
					CT3001	Implantable metal	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	08/01/2023	08/01/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	08/01/2023	08/01/2023
					CT3002	Implantable synthetic polymer	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	08/01/2023	08/01/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	08/01/2023	08/01/2023
					CT3006	Implantable inorganic material	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	08/01/2023	08/01/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	08/01/2023	08/01/2023
					CT314	Electrically powered	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	08/01/2023	08/01/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	08/01/2023	08/01/2023
					CT321	Long-term surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	08/01/2023	08/01/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	08/01/2023	08/01/2023
					CT449	Generators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	08/01/2023	08/01/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	08/01/2023	08/01/2023
					CT981	Single use	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	08/01/2023	08/01/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	08/01/2023	08/01/2023
					CT983	Surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	08/01/2023	08/01/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	08/01/2023	08/01/2023
12226	2023		47265	Dual-chamber implantable pacemaker, rate-responsive	CT141	Electrodes/Leads and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	10/01/2023	10/01/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	10/01/2023	10/01/2023
					CT145	Cardiovascular devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	10/01/2023	10/01/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	10/01/2023	10/01/2023
					CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	10/01/2023	10/01/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	10/01/2023	10/01/2023
					CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	10/01/2023	10/01/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	10/01/2023	10/01/2023
					CT233	Surgical	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	10/01/2023	10/01/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	10/01/2023	10/01/2023
					CT2406	Implantable	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	10/01/2023	10/01/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	10/01/2023	10/01/2023
					CT2407	Active implantable	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	10/01/2023	10/01/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	10/01/2023	10/01/2023
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	10/01/2023	10/01/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	10/01/2023	10/01/2023
					CT2727	Active implant devices/ancillaries	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	10/01/2023	10/01/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	10/01/2023	10/01/2023
					CT2750	Implantable cardiac electrical stimulators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	10/01/2023	10/01/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	10/01/2023	10/01/2023
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	10/01/2023	10/01/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	10/01/2023	10/01/2023
					CT2985	Active	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	10/01/2023	10/01/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	10/01/2023	10/01/2023
					CT3001	Implantable metal	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	10/01/2023	10/01/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	10/01/2023	10/01/2023
					CT3002	Implantable synthetic polymer	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	10/01/2023	10/01/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	10/01/2023	10/01/2023
					CT3006	Implantable inorganic material	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	10/01/2023	10/01/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	10/01/2023	10/01/2023
					CT314	Electrically powered	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	10/01/2023	10/01/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	10/01/2023	10/01/2023
					CT321	Long-term surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	10/01/2023	10/01/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	10/01/2023	10/01/2023
					CT449	Generators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	10/01/2023	10/01/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	10/01/2023	10/01/2023
					CT981	Single use	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	10/01/2023	10/01/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	10/01/2023	10/01/2023
					CT983	Surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	10/01/2023	10/01/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	10/01/2023	10/01/2023
12227	2023		47265	Dual-chamber implantable pacemaker, rate-responsive	CT141	Electrodes/Leads and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	13/01/2023	13/01/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	13/01/2023	13/01/2023
					CT145	Cardiovascular devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	13/01/2023	13/01/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	13/01/2023	13/01/2023
					CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	13/01/2023	13/01/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	13/01/2023	13/01/2023
					CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	13/01/2023	13/01/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	13/01/2023	13/01/2023

					CT233	Surgical	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	13/01/2023	13/01/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	13/01/2023	13/01/2023
					CT2406	Implantable	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	13/01/2023	13/01/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	13/01/2023	13/01/2023
					CT2407	Active implantable	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	13/01/2023	13/01/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	13/01/2023	13/01/2023
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	13/01/2023	13/01/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	13/01/2023	13/01/2023
					CT2727	Active implant devices/ancillaries	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	13/01/2023	13/01/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	13/01/2023	13/01/2023
					CT2750	Implantable cardiac electrical stimulators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	13/01/2023	13/01/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	13/01/2023	13/01/2023
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	13/01/2023	13/01/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	13/01/2023	13/01/2023
					CT2985	Active	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	13/01/2023	13/01/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	13/01/2023	13/01/2023
					CT3001	Implantable metal	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	13/01/2023	13/01/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	13/01/2023	13/01/2023
					CT3002	Implantable synthetic polymer	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	13/01/2023	13/01/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	13/01/2023	13/01/2023
					CT3006	Implantable inorganic material	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	13/01/2023	13/01/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	13/01/2023	13/01/2023
					CT314	Electrically powered	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	13/01/2023	13/01/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	13/01/2023	13/01/2023
					CT321	Long-term surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	13/01/2023	13/01/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	13/01/2023	13/01/2023
					CT449	Generators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	13/01/2023	13/01/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	13/01/2023	13/01/2023
					CT981	Single use	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	13/01/2023	13/01/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	13/01/2023	13/01/2023
					CT983	Surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	13/01/2023	13/01/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	13/01/2023	13/01/2023
12228	2023		47265	Dual-chamber implantable pacemaker, rate-responsive	CT141	Electrodes/Leads and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	15/01/2023	15/01/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	15/01/2023	15/01/2023
					CT145	Cardiovascular devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	15/01/2023	15/01/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	15/01/2023	15/01/2023
					CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	15/01/2023	15/01/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	15/01/2023	15/01/2023
					CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	15/01/2023	15/01/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	15/01/2023	15/01/2023
					CT233	Surgical	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	15/01/2023	15/01/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	15/01/2023	15/01/2023
					CT2406	Implantable	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	15/01/2023	15/01/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	15/01/2023	15/01/2023
					CT2407	Active implantable	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	15/01/2023	15/01/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	15/01/2023	15/01/2023
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	15/01/2023	15/01/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	15/01/2023	15/01/2023
					CT2727	Active implant devices/ancillaries	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	15/01/2023	15/01/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	15/01/2023	15/01/2023
					CT2750	Implantable cardiac electrical stimulators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	15/01/2023	15/01/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	15/01/2023	15/01/2023
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	15/01/2023	15/01/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	15/01/2023	15/01/2023
					CT2985	Active	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	15/01/2023	15/01/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	15/01/2023	15/01/2023
					CT3001	Implantable metal	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	15/01/2023	15/01/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	15/01/2023	15/01/2023
					CT3002	Implantable synthetic polymer	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	15/01/2023	15/01/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	15/01/2023	15/01/2023
					CT3006	Implantable inorganic material	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	15/01/2023	15/01/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	15/01/2023	15/01/2023
					CT314	Electrically powered	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	15/01/2023	15/01/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	15/01/2023	15/01/2023
					CT321	Long-term surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	15/01/2023	15/01/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	15/01/2023	15/01/2023
					CT449	Generators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	15/01/2023	15/01/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	15/01/2023	15/01/2023
					CT981	Single use	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	15/01/2023	15/01/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	15/01/2023	15/01/2023
					CT983	Surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	15/01/2023	15/01/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	15/01/2023	15/01/2023
12229	2023		47265	Dual-chamber implantable pacemaker, rate-responsive	CT141	Electrodes/Leads and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	15/01/2023	15/01/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	15/01/2023	15/01/2023
					CT145	Cardiovascular devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	15/01/2023	15/01/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	15/01/2023	15/01/2023
					CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	15/01/2023	15/01/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	15/01/2023	15/01/2023
					CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	15/01/2023	15/01/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	15/01/2023	15/01/2023
					CT233	Surgical	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	15/01/2023	15/01/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	15/01/2023	15/01/2023
					CT2406	Implantable	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	15/01/2023	15/01/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	15/01/2023	15/01/2023
					CT2407	Active implantable	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	15/01/2023	15/01/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	15/01/2023	15/01/2023
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	15/01/2023	15/01/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	15/01/2023	15/01/2023

					CT2727	Active implant devices/ancillaries	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	15/01/2023	15/01/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	15/01/2023	15/01/2023
					CT2750	Implantable cardiac electrical stimulators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	15/01/2023	15/01/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	15/01/2023	15/01/2023
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	15/01/2023	15/01/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	15/01/2023	15/01/2023
					CT2985	Active	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	15/01/2023	15/01/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	15/01/2023	15/01/2023
					CT3001	Implantable metal	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	15/01/2023	15/01/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	15/01/2023	15/01/2023
					CT3002	Implantable synthetic polymer	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	15/01/2023	15/01/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	15/01/2023	15/01/2023
					CT3006	Implantable inorganic material	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	15/01/2023	15/01/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	15/01/2023	15/01/2023
					CT314	Electrically powered	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	15/01/2023	15/01/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	15/01/2023	15/01/2023
					CT321	Long-term surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	15/01/2023	15/01/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	15/01/2023	15/01/2023
					CT449	Generators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	15/01/2023	15/01/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	15/01/2023	15/01/2023
					CT981	Single use	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	15/01/2023	15/01/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	15/01/2023	15/01/2023
					CT983	Surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	15/01/2023	15/01/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	15/01/2023	15/01/2023
12230	2023		47265	Dual-chamber implantable pacemaker, rate-responsive	CT141	Electrodes/Leads and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	15/01/2023	15/01/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	15/01/2023	15/01/2023
					CT145	Cardiovascular devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	15/01/2023	15/01/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	15/01/2023	15/01/2023
					CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	15/01/2023	15/01/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	15/01/2023	15/01/2023
					CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	15/01/2023	15/01/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	15/01/2023	15/01/2023
					CT233	Surgical	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	15/01/2023	15/01/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	15/01/2023	15/01/2023
					CT2406	Implantable	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	15/01/2023	15/01/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	15/01/2023	15/01/2023
					CT2407	Active implantable	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	15/01/2023	15/01/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	15/01/2023	15/01/2023
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	15/01/2023	15/01/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	15/01/2023	15/01/2023
					CT2727	Active implant devices/ancillaries	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	15/01/2023	15/01/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	15/01/2023	15/01/2023
					CT2750	Implantable cardiac electrical stimulators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	15/01/2023	15/01/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	15/01/2023	15/01/2023
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	15/01/2023	15/01/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	15/01/2023	15/01/2023
					CT2985	Active	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	15/01/2023	15/01/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	15/01/2023	15/01/2023
					CT3001	Implantable metal	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	15/01/2023	15/01/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	15/01/2023	15/01/2023
					CT3002	Implantable synthetic polymer	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	15/01/2023	15/01/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	15/01/2023	15/01/2023
					CT3006	Implantable inorganic material	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	15/01/2023	15/01/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	15/01/2023	15/01/2023
					CT314	Electrically powered	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	15/01/2023	15/01/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	15/01/2023	15/01/2023
					CT321	Long-term surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	15/01/2023	15/01/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	15/01/2023	15/01/2023
					CT449	Generators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	15/01/2023	15/01/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	15/01/2023	15/01/2023
					CT981	Single use	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	15/01/2023	15/01/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	15/01/2023	15/01/2023
					CT983	Surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	15/01/2023	15/01/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	15/01/2023	15/01/2023
12231	2023		47265	Dual-chamber implantable pacemaker, rate-responsive	CT141	Electrodes/Leads and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	15/01/2023	15/01/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	15/01/2023	15/01/2023
					CT145	Cardiovascular devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	15/01/2023	15/01/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	15/01/2023	15/01/2023
					CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	15/01/2023	15/01/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	15/01/2023	15/01/2023
					CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	15/01/2023	15/01/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	15/01/2023	15/01/2023
					CT233	Surgical	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	15/01/2023	15/01/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	15/01/2023	15/01/2023
					CT2406	Implantable	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	15/01/2023	15/01/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	15/01/2023	15/01/2023
					CT2407	Active implantable	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	15/01/2023	15/01/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	15/01/2023	15/01/2023
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	15/01/2023	15/01/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	15/01/2023	15/01/2023
					CT2727	Active implant devices/ancillaries	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	15/01/2023	15/01/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	15/01/2023	15/01/2023
					CT2750	Implantable cardiac electrical stimulators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	15/01/2023	15/01/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	15/01/2023	15/01/2023
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	15/01/2023	15/01/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	15/01/2023	15/01/2023
					CT2985	Active	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	15/01/2023	15/01/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	15/01/2023	15/01/2023

					CT3001	Implantable metal	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	15/01/2023	15/01/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	15/01/2023	15/01/2023
					CT3002	Implantable synthetic polymer	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	15/01/2023	15/01/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	15/01/2023	15/01/2023
					CT3006	Implantable inorganic material	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	15/01/2023	15/01/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	15/01/2023	15/01/2023
					CT314	Electrically powered	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	15/01/2023	15/01/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	15/01/2023	15/01/2023
					CT321	Long-term surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	15/01/2023	15/01/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	15/01/2023	15/01/2023
					CT449	Generators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	15/01/2023	15/01/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	15/01/2023	15/01/2023
					CT981	Single use	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	15/01/2023	15/01/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	15/01/2023	15/01/2023
					CT983	Surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	15/01/2023	15/01/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	15/01/2023	15/01/2023
12232	2023	All other reportable incidents	47270	Cardiac resynchronization therapy implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E2104	Electric Shock	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
									A071302	Inappropriate/Inadequate Shock/Stimulation		
							F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
									A071302	Inappropriate/Inadequate Shock/Stimulation		
12233	2023	All other reportable incidents	36241	Pacemaker/defibrillator lead	CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A072201	High Impedance		
							F26	No Health Consequences or Impact	A072201	High Impedance		
12234	2023	All other reportable incidents	37265	Dual-chamber implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A0713	Defibrillation/Stimulation Problem		
									A0722	Impedance Problem		
									A090804	False Positive Result		
									A090805	Incorrect Measurement		
							F26	No Health Consequences or Impact	A0713	Defibrillation/Stimulation Problem		
									A0722	Impedance Problem		
									A090804	False Positive Result		
									A090805	Incorrect Measurement		
12235	2023		35224	External pacemaker, epicardial pacing	CT1684	Cardiac pacemakers and associated devices	F27	Problem Identified During Non-Clinical Procedure	A0201	Product Quality Problem		
					CT1685	Cardiac pacemakers	F27	Problem Identified During Non-Clinical Procedure	A0201	Product Quality Problem		
12236	2023	Death	17882	Physiologic-monitoring defibrillation system	CT1684	Cardiac pacemakers and associated devices	E2401	Insufficient Information	A23	Use of Device Problem		
							F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A23	Use of Device Problem		
							F24	Insufficient Health Impact Information	A23	Use of Device Problem		
12237	2023		47265	Dual-chamber implantable pacemaker, rate-responsive	CT141	Electrodes/Leads and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	19/01/2023	19/01/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	19/01/2023	19/01/2023
					CT145	Cardiovascular devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	19/01/2023	19/01/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	19/01/2023	19/01/2023
					CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	19/01/2023	19/01/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	19/01/2023	19/01/2023
					CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	19/01/2023	19/01/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	19/01/2023	19/01/2023
					CT233	Surgical	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	19/01/2023	19/01/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	19/01/2023	19/01/2023
					CT2406	Implantable	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	19/01/2023	19/01/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	19/01/2023	19/01/2023
					CT2407	Active implantable	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	19/01/2023	19/01/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	19/01/2023	19/01/2023
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	19/01/2023	19/01/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	19/01/2023	19/01/2023
					CT2727	Active implant devices/ancillaries	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	19/01/2023	19/01/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	19/01/2023	19/01/2023
					CT2750	Implantable cardiac electrical stimulators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	19/01/2023	19/01/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	19/01/2023	19/01/2023
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	19/01/2023	19/01/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	19/01/2023	19/01/2023
					CT2985	Active	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	19/01/2023	19/01/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	19/01/2023	19/01/2023
					CT3001	Implantable metal	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	19/01/2023	19/01/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	19/01/2023	19/01/2023
					CT3002	Implantable synthetic polymer	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	19/01/2023	19/01/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	19/01/2023	19/01/2023
					CT3006	Implantable inorganic material	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	19/01/2023	19/01/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	19/01/2023	19/01/2023
					CT314	Electrically powered	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	19/01/2023	19/01/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	19/01/2023	19/01/2023
					CT321	Long-term surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	19/01/2023	19/01/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	19/01/2023	19/01/2023
					CT449	Generators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	19/01/2023	19/01/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	19/01/2023	19/01/2023
					CT981	Single use	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	19/01/2023	19/01/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	19/01/2023	19/01/2023
					CT983	Surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	19/01/2023	19/01/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	19/01/2023	19/01/2023
12238	2023	All other reportable incidents	47265	Dual-chamber implantable pacemaker, rate-responsive	CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem	26/04/2023	26/04/2023
									A070504	Premature Discharge of Battery	26/04/2023	26/04/2023
							F08	Hospitalization or Prolonged Hospitalization	A0705	Battery Problem	26/04/2023	26/04/2023
									A070504	Premature Discharge of Battery	26/04/2023	26/04/2023
							F1905	Device Revision or Replacement	A0705	Battery Problem	26/04/2023	26/04/2023
									A070504	Premature Discharge of Battery	26/04/2023	26/04/2023
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem	26/04/2023	26/04/2023
									A070504	Premature Discharge of Battery	26/04/2023	26/04/2023
							F08	Hospitalization or Prolonged Hospitalization	A0705	Battery Problem	26/04/2023	26/04/2023
									A070504	Premature Discharge of Battery	26/04/2023	26/04/2023
							F1905	Device Revision or Replacement	A0705	Battery Problem	26/04/2023	26/04/2023

								A070504	Premature Discharge of Battery	26/04/2023	26/04/2023
				CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem	26/04/2023	26/04/2023
								A070504	Premature Discharge of Battery	26/04/2023	26/04/2023
						F08	Hospitalization or Prolonged Hospitalization	A0705	Battery Problem	26/04/2023	26/04/2023
								A070504	Premature Discharge of Battery	26/04/2023	26/04/2023
						F1905	Device Revision or Replacement	A0705	Battery Problem	26/04/2023	26/04/2023
								A070504	Premature Discharge of Battery	26/04/2023	26/04/2023
12239	2023	All other reportable incidents	47267	Single-chamber implantable pacemaker, rate-responsive	CT141	Electrodes/Leads and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	01/02/2023	01/02/2023
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	01/02/2023	01/02/2023
							F1903	Device Explantation	A070504	Premature Discharge of Battery	01/02/2023	01/02/2023
					CT145	Cardiovascular devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	01/02/2023	01/02/2023
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	01/02/2023	01/02/2023
							F1903	Device Explantation	A070504	Premature Discharge of Battery	01/02/2023	01/02/2023
					CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	01/02/2023	01/02/2023
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	01/02/2023	01/02/2023
							F1903	Device Explantation	A070504	Premature Discharge of Battery	01/02/2023	01/02/2023
					CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	01/02/2023	01/02/2023
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	01/02/2023	01/02/2023
							F1903	Device Explantation	A070504	Premature Discharge of Battery	01/02/2023	01/02/2023
					CT233	Surgical	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	01/02/2023	01/02/2023
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	01/02/2023	01/02/2023
							F1903	Device Explantation	A070504	Premature Discharge of Battery	01/02/2023	01/02/2023
					CT2406	Implantable	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	01/02/2023	01/02/2023
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	01/02/2023	01/02/2023
							F1903	Device Explantation	A070504	Premature Discharge of Battery	01/02/2023	01/02/2023
					CT2407	Active implantable	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	01/02/2023	01/02/2023
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	01/02/2023	01/02/2023
							F1903	Device Explantation	A070504	Premature Discharge of Battery	01/02/2023	01/02/2023
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	01/02/2023	01/02/2023
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	01/02/2023	01/02/2023
							F1903	Device Explantation	A070504	Premature Discharge of Battery	01/02/2023	01/02/2023
					CT2727	Active implant devices/ancillaries	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	01/02/2023	01/02/2023
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	01/02/2023	01/02/2023
							F1903	Device Explantation	A070504	Premature Discharge of Battery	01/02/2023	01/02/2023
					CT2750	Implantable cardiac electrical stimulators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	01/02/2023	01/02/2023
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	01/02/2023	01/02/2023
							F1903	Device Explantation	A070504	Premature Discharge of Battery	01/02/2023	01/02/2023
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	01/02/2023	01/02/2023
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	01/02/2023	01/02/2023
							F1903	Device Explantation	A070504	Premature Discharge of Battery	01/02/2023	01/02/2023
					CT2985	Active	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	01/02/2023	01/02/2023
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	01/02/2023	01/02/2023
							F1903	Device Explantation	A070504	Premature Discharge of Battery	01/02/2023	01/02/2023
					CT3001	Implantable metal	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	01/02/2023	01/02/2023
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	01/02/2023	01/02/2023
							F1903	Device Explantation	A070504	Premature Discharge of Battery	01/02/2023	01/02/2023
					CT3002	Implantable synthetic polymer	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	01/02/2023	01/02/2023
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	01/02/2023	01/02/2023
							F1903	Device Explantation	A070504	Premature Discharge of Battery	01/02/2023	01/02/2023
					CT3006	Implantable inorganic material	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	01/02/2023	01/02/2023
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	01/02/2023	01/02/2023
							F1903	Device Explantation	A070504	Premature Discharge of Battery	01/02/2023	01/02/2023
					CT314	Electrically powered	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	01/02/2023	01/02/2023
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	01/02/2023	01/02/2023
							F1903	Device Explantation	A070504	Premature Discharge of Battery	01/02/2023	01/02/2023
					CT321	Long-term surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	01/02/2023	01/02/2023
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	01/02/2023	01/02/2023
							F1903	Device Explantation	A070504	Premature Discharge of Battery	01/02/2023	01/02/2023
					CT449	Generators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	01/02/2023	01/02/2023
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	01/02/2023	01/02/2023
							F1903	Device Explantation	A070504	Premature Discharge of Battery	01/02/2023	01/02/2023
					CT981	Single use	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	01/02/2023	01/02/2023
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	01/02/2023	01/02/2023
							F1903	Device Explantation	A070504	Premature Discharge of Battery	01/02/2023	01/02/2023
					CT983	Surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	01/02/2023	01/02/2023
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	01/02/2023	01/02/2023
							F1903	Device Explantation	A070504	Premature Discharge of Battery	01/02/2023	01/02/2023
12240	2023	All other reportable incidents	47265	Dual-chamber implantable pacemaker, rate-responsive	CT141	Electrodes/Leads and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	29/11/2022	29/11/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	29/11/2022	29/11/2022
					CT145	Cardiovascular devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	29/11/2022	29/11/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	29/11/2022	29/11/2022
					CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	29/11/2022	29/11/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	29/11/2022	29/11/2022
					CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	29/11/2022	29/11/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	29/11/2022	29/11/2022
					CT233	Surgical	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	29/11/2022	29/11/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	29/11/2022	29/11/2022
					CT2406	Implantable	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	29/11/2022	29/11/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	29/11/2022	29/11/2022
					CT2407	Active implantable	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	29/11/2022	29/11/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	29/11/2022	29/11/2022
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	29/11/2022	29/11/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	29/11/2022	29/11/2022
					CT2727	Active implant devices/ancillaries	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	29/11/2022	29/11/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	29/11/2022	29/11/2022
					CT2750	Implantable cardiac electrical stimulators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	29/11/2022	29/11/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	29/11/2022	29/11/2022
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	29/11/2022	29/11/2022

							F1901	Additional Surgery	A24	No Apparent Device Problem	29/11/2022	29/11/2022
					CT2985	Active	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	29/11/2022	29/11/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	29/11/2022	29/11/2022
					CT3001	Implantable metal	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	29/11/2022	29/11/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	29/11/2022	29/11/2022
					CT3002	Implantable synthetic polymer	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	29/11/2022	29/11/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	29/11/2022	29/11/2022
					CT3006	Implantable inorganic material	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	29/11/2022	29/11/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	29/11/2022	29/11/2022
					CT314	Electrically powered	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	29/11/2022	29/11/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	29/11/2022	29/11/2022
					CT321	Long-term surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	29/11/2022	29/11/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	29/11/2022	29/11/2022
					CT449	Generators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	29/11/2022	29/11/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	29/11/2022	29/11/2022
					CT981	Single use	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	29/11/2022	29/11/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	29/11/2022	29/11/2022
					CT983	Surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	29/11/2022	29/11/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	29/11/2022	29/11/2022
12241	2023		47265	Dual-chamber implantable pacemaker, rate-responsive	CT141	Electrodes/Leads and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem		
							F26	No Health Consequences or Impact	A24	No Apparent Device Problem		
					CT145	Cardiovascular devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem		
							F26	No Health Consequences or Impact	A24	No Apparent Device Problem		
					CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem		
							F26	No Health Consequences or Impact	A24	No Apparent Device Problem		
					CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem		
							F26	No Health Consequences or Impact	A24	No Apparent Device Problem		
					CT233	Surgical	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem		
							F26	No Health Consequences or Impact	A24	No Apparent Device Problem		
					CT2406	Implantable	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem		
							F26	No Health Consequences or Impact	A24	No Apparent Device Problem		
					CT2407	Active implantable	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem		
							F26	No Health Consequences or Impact	A24	No Apparent Device Problem		
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem		
							F26	No Health Consequences or Impact	A24	No Apparent Device Problem		
					CT2727	Active implant devices/ancillaries	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem		
							F26	No Health Consequences or Impact	A24	No Apparent Device Problem		
					CT2750	Implantable cardiac electrical stimulators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem		
							F26	No Health Consequences or Impact	A24	No Apparent Device Problem		
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem		
							F26	No Health Consequences or Impact	A24	No Apparent Device Problem		
					CT2985	Active	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem		
							F26	No Health Consequences or Impact	A24	No Apparent Device Problem		
					CT3001	Implantable metal	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem		
							F26	No Health Consequences or Impact	A24	No Apparent Device Problem		
					CT3002	Implantable synthetic polymer	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem		
							F26	No Health Consequences or Impact	A24	No Apparent Device Problem		
					CT3006	Implantable inorganic material	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem		
							F26	No Health Consequences or Impact	A24	No Apparent Device Problem		
					CT314	Electrically powered	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem		
							F26	No Health Consequences or Impact	A24	No Apparent Device Problem		
					CT321	Long-term surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem		
							F26	No Health Consequences or Impact	A24	No Apparent Device Problem		
					CT449	Generators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem		
							F26	No Health Consequences or Impact	A24	No Apparent Device Problem		
					CT981	Single use	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem		
							F26	No Health Consequences or Impact	A24	No Apparent Device Problem		
					CT983	Surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem		
							F26	No Health Consequences or Impact	A24	No Apparent Device Problem		
12242	2023		47265	Dual-chamber implantable pacemaker, rate-responsive	CT141	Electrodes/Leads and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	03/02/2023	03/02/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	03/02/2023	03/02/2023
					CT145	Cardiovascular devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	03/02/2023	03/02/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	03/02/2023	03/02/2023
					CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	03/02/2023	03/02/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	03/02/2023	03/02/2023
					CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	03/02/2023	03/02/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	03/02/2023	03/02/2023
					CT233	Surgical	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	03/02/2023	03/02/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	03/02/2023	03/02/2023
					CT2406	Implantable	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	03/02/2023	03/02/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	03/02/2023	03/02/2023
					CT2407	Active implantable	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	03/02/2023	03/02/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	03/02/2023	03/02/2023
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	03/02/2023	03/02/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	03/02/2023	03/02/2023
					CT2727	Active implant devices/ancillaries	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	03/02/2023	03/02/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	03/02/2023	03/02/2023
					CT2750	Implantable cardiac electrical stimulators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	03/02/2023	03/02/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	03/02/2023	03/02/2023
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	03/02/2023	03/02/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	03/02/2023	03/02/2023
					CT2985	Active	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	03/02/2023	03/02/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	03/02/2023	03/02/2023
					CT3001	Implantable metal	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	03/02/2023	03/02/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	03/02/2023	03/02/2023
					CT3002	Implantable synthetic polymer	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	03/02/2023	03/02/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	03/02/2023	03/02/2023
					CT3006	Implantable inorganic material	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	03/02/2023	03/02/2023

							F1901	Additional Surgery	A24	No Apparent Device Problem	03/02/2023	03/02/2023
					CT314	Electrically powered	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	03/02/2023	03/02/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	03/02/2023	03/02/2023
					CT321	Long-term surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	03/02/2023	03/02/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	03/02/2023	03/02/2023
					CT449	Generators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	03/02/2023	03/02/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	03/02/2023	03/02/2023
					CT981	Single use	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	03/02/2023	03/02/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	03/02/2023	03/02/2023
					CT983	Surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	03/02/2023	03/02/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	03/02/2023	03/02/2023
12243	2023	All other reportable incidents	35852	Single-chamber implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
									A0908	Incorrect, Inadequate or Imprecise Result or Readings		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
									A0908	Incorrect, Inadequate or Imprecise Result or Readings		
12244	2023	All other reportable incidents	47265	Dual-chamber implantable pacemaker, rate-responsive	CT141	Electrodes/Leads and associated devices	E011903	Syncope/Fainting	A070103	Intermittent Capture	25/01/2023	25/01/2023
							F1905	Device Revision or Replacement	A070103	Intermittent Capture	25/01/2023	25/01/2023
					CT145	Cardiovascular devices	E011903	Syncope/Fainting	A070103	Intermittent Capture	25/01/2023	25/01/2023
							F1905	Device Revision or Replacement	A070103	Intermittent Capture	25/01/2023	25/01/2023
					CT1684	Cardiac pacemakers and associated devices	E011903	Syncope/Fainting	A070103	Intermittent Capture	25/01/2023	25/01/2023
							F1905	Device Revision or Replacement	A070103	Intermittent Capture	25/01/2023	25/01/2023
					CT1685	Cardiac pacemakers	E011903	Syncope/Fainting	A070103	Intermittent Capture	25/01/2023	25/01/2023
							F1905	Device Revision or Replacement	A070103	Intermittent Capture	25/01/2023	25/01/2023
					CT233	Surgical	E011903	Syncope/Fainting	A070103	Intermittent Capture	25/01/2023	25/01/2023
							F1905	Device Revision or Replacement	A070103	Intermittent Capture	25/01/2023	25/01/2023
					CT2406	Implantable	E011903	Syncope/Fainting	A070103	Intermittent Capture	25/01/2023	25/01/2023
							F1905	Device Revision or Replacement	A070103	Intermittent Capture	25/01/2023	25/01/2023
					CT2407	Active implantable	E011903	Syncope/Fainting	A070103	Intermittent Capture	25/01/2023	25/01/2023
							F1905	Device Revision or Replacement	A070103	Intermittent Capture	25/01/2023	25/01/2023
					CT257	Implantable cardiac pacemakers	E011903	Syncope/Fainting	A070103	Intermittent Capture	25/01/2023	25/01/2023
							F1905	Device Revision or Replacement	A070103	Intermittent Capture	25/01/2023	25/01/2023
					CT2727	Active implant devices/ancillaries	E011903	Syncope/Fainting	A070103	Intermittent Capture	25/01/2023	25/01/2023
							F1905	Device Revision or Replacement	A070103	Intermittent Capture	25/01/2023	25/01/2023
					CT2750	Implantable cardiac electrical stimulators and associated devices	E011903	Syncope/Fainting	A070103	Intermittent Capture	25/01/2023	25/01/2023
							F1905	Device Revision or Replacement	A070103	Intermittent Capture	25/01/2023	25/01/2023
					CT2866	Implantable cardiac electrical stimulators	E011903	Syncope/Fainting	A070103	Intermittent Capture	25/01/2023	25/01/2023
							F1905	Device Revision or Replacement	A070103	Intermittent Capture	25/01/2023	25/01/2023
					CT2985	Active	E011903	Syncope/Fainting	A070103	Intermittent Capture	25/01/2023	25/01/2023
							F1905	Device Revision or Replacement	A070103	Intermittent Capture	25/01/2023	25/01/2023
					CT3001	Implantable metal	E011903	Syncope/Fainting	A070103	Intermittent Capture	25/01/2023	25/01/2023
							F1905	Device Revision or Replacement	A070103	Intermittent Capture	25/01/2023	25/01/2023
					CT3002	Implantable synthetic polymer	E011903	Syncope/Fainting	A070103	Intermittent Capture	25/01/2023	25/01/2023
							F1905	Device Revision or Replacement	A070103	Intermittent Capture	25/01/2023	25/01/2023
					CT3006	Implantable inorganic material	E011903	Syncope/Fainting	A070103	Intermittent Capture	25/01/2023	25/01/2023
							F1905	Device Revision or Replacement	A070103	Intermittent Capture	25/01/2023	25/01/2023
					CT314	Electrically powered	E011903	Syncope/Fainting	A070103	Intermittent Capture	25/01/2023	25/01/2023
							F1905	Device Revision or Replacement	A070103	Intermittent Capture	25/01/2023	25/01/2023
					CT321	Long-term surgical invasive	E011903	Syncope/Fainting	A070103	Intermittent Capture	25/01/2023	25/01/2023
							F1905	Device Revision or Replacement	A070103	Intermittent Capture	25/01/2023	25/01/2023
					CT449	Generators and associated devices	E011903	Syncope/Fainting	A070103	Intermittent Capture	25/01/2023	25/01/2023
							F1905	Device Revision or Replacement	A070103	Intermittent Capture	25/01/2023	25/01/2023
					CT981	Single use	E011903	Syncope/Fainting	A070103	Intermittent Capture	25/01/2023	25/01/2023
							F1905	Device Revision or Replacement	A070103	Intermittent Capture	25/01/2023	25/01/2023
					CT983	Surgical invasive	E011903	Syncope/Fainting	A070103	Intermittent Capture	25/01/2023	25/01/2023
							F1905	Device Revision or Replacement	A070103	Intermittent Capture	25/01/2023	25/01/2023
12245	2023		47265	Dual-chamber implantable pacemaker, rate-responsive	CT141	Electrodes/Leads and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	25/01/2023	25/01/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	25/01/2023	25/01/2023
					CT145	Cardiovascular devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	25/01/2023	25/01/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	25/01/2023	25/01/2023
					CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	25/01/2023	25/01/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	25/01/2023	25/01/2023
					CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	25/01/2023	25/01/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	25/01/2023	25/01/2023
					CT233	Surgical	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	25/01/2023	25/01/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	25/01/2023	25/01/2023
					CT2406	Implantable	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	25/01/2023	25/01/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	25/01/2023	25/01/2023
					CT2407	Active implantable	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	25/01/2023	25/01/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	25/01/2023	25/01/2023
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	25/01/2023	25/01/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	25/01/2023	25/01/2023
					CT2727	Active implant devices/ancillaries	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	25/01/2023	25/01/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	25/01/2023	25/01/2023
					CT2750	Implantable cardiac electrical stimulators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	25/01/2023	25/01/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	25/01/2023	25/01/2023
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	25/01/2023	25/01/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	25/01/2023	25/01/2023
					CT2985	Active	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	25/01/2023	25/01/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	25/01/2023	25/01/2023
					CT3001	Implantable metal	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	25/01/2023	25/01/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	25/01/2023	25/01/2023
					CT3002	Implantable synthetic polymer	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	25/01/2023	25/01/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	25/01/2023	25/01/2023
					CT3006	Implantable inorganic material	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	25/01/2023	25/01/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	25/01/2023	25/01/2023
					CT314	Electrically powered	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	25/01/2023	25/01/2023

							F1901	Additional Surgery	A24	No Apparent Device Problem	25/01/2023	25/01/2023
					CT321	Long-term surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	25/01/2023	25/01/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	25/01/2023	25/01/2023
					CT449	Generators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	25/01/2023	25/01/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	25/01/2023	25/01/2023
					CT981	Single use	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	25/01/2023	25/01/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	25/01/2023	25/01/2023
					CT983	Surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	25/01/2023	25/01/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	25/01/2023	25/01/2023
12246	2023	All other reportable incidents	35223	Endocardial/interventricular septal pacing lead	CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070102	High Capture Threshold		
									A070104	Unstable Capture Threshold		
									A0722	Impedance Problem		
									A072202	Low Impedance		
									A090801	Signal Artifact/Noise		
							F23	Unexpected Medical Intervention	A070102	High Capture Threshold		
									A070104	Unstable Capture Threshold		
									A0722	Impedance Problem		
									A072202	Low Impedance		
									A090801	Signal Artifact/Noise		
12247	2023	All other reportable incidents	47265	Dual-chamber implantable pacemaker, rate-responsive	CT141	Electrodes/Leads and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A090804	False Positive Result	26/01/2023	26/01/2023
							F1905	Device Revision or Replacement	A090804	False Positive Result	26/01/2023	26/01/2023
					CT145	Cardiovascular devices	E2403	No Clinical Signs, Symptoms or Conditions	A090804	False Positive Result	26/01/2023	26/01/2023
							F1905	Device Revision or Replacement	A090804	False Positive Result	26/01/2023	26/01/2023
					CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A090804	False Positive Result	26/01/2023	26/01/2023
							F1905	Device Revision or Replacement	A090804	False Positive Result	26/01/2023	26/01/2023
					CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A090804	False Positive Result	26/01/2023	26/01/2023
							F1905	Device Revision or Replacement	A090804	False Positive Result	26/01/2023	26/01/2023
					CT233	Surgical	E2403	No Clinical Signs, Symptoms or Conditions	A090804	False Positive Result	26/01/2023	26/01/2023
							F1905	Device Revision or Replacement	A090804	False Positive Result	26/01/2023	26/01/2023
					CT2406	Implantable	E2403	No Clinical Signs, Symptoms or Conditions	A090804	False Positive Result	26/01/2023	26/01/2023
							F1905	Device Revision or Replacement	A090804	False Positive Result	26/01/2023	26/01/2023
					CT2407	Active implantable	E2403	No Clinical Signs, Symptoms or Conditions	A090804	False Positive Result	26/01/2023	26/01/2023
							F1905	Device Revision or Replacement	A090804	False Positive Result	26/01/2023	26/01/2023
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A090804	False Positive Result	26/01/2023	26/01/2023
							F1905	Device Revision or Replacement	A090804	False Positive Result	26/01/2023	26/01/2023
					CT2727	Active implant devices/ancillaries	E2403	No Clinical Signs, Symptoms or Conditions	A090804	False Positive Result	26/01/2023	26/01/2023
							F1905	Device Revision or Replacement	A090804	False Positive Result	26/01/2023	26/01/2023
					CT2750	Implantable cardiac electrical stimulators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A090804	False Positive Result	26/01/2023	26/01/2023
							F1905	Device Revision or Replacement	A090804	False Positive Result	26/01/2023	26/01/2023
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A090804	False Positive Result	26/01/2023	26/01/2023
							F1905	Device Revision or Replacement	A090804	False Positive Result	26/01/2023	26/01/2023
					CT2985	Active	E2403	No Clinical Signs, Symptoms or Conditions	A090804	False Positive Result	26/01/2023	26/01/2023
							F1905	Device Revision or Replacement	A090804	False Positive Result	26/01/2023	26/01/2023
					CT3001	Implantable metal	E2403	No Clinical Signs, Symptoms or Conditions	A090804	False Positive Result	26/01/2023	26/01/2023
							F1905	Device Revision or Replacement	A090804	False Positive Result	26/01/2023	26/01/2023
					CT3002	Implantable synthetic polymer	E2403	No Clinical Signs, Symptoms or Conditions	A090804	False Positive Result	26/01/2023	26/01/2023
							F1905	Device Revision or Replacement	A090804	False Positive Result	26/01/2023	26/01/2023
					CT3006	Implantable inorganic material	E2403	No Clinical Signs, Symptoms or Conditions	A090804	False Positive Result	26/01/2023	26/01/2023
							F1905	Device Revision or Replacement	A090804	False Positive Result	26/01/2023	26/01/2023
					CT314	Electrically powered	E2403	No Clinical Signs, Symptoms or Conditions	A090804	False Positive Result	26/01/2023	26/01/2023
							F1905	Device Revision or Replacement	A090804	False Positive Result	26/01/2023	26/01/2023
					CT321	Long-term surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A090804	False Positive Result	26/01/2023	26/01/2023
							F1905	Device Revision or Replacement	A090804	False Positive Result	26/01/2023	26/01/2023
					CT449	Generators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A090804	False Positive Result	26/01/2023	26/01/2023
							F1905	Device Revision or Replacement	A090804	False Positive Result	26/01/2023	26/01/2023
					CT981	Single use	E2403	No Clinical Signs, Symptoms or Conditions	A090804	False Positive Result	26/01/2023	26/01/2023
							F1905	Device Revision or Replacement	A090804	False Positive Result	26/01/2023	26/01/2023
					CT983	Surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A090804	False Positive Result	26/01/2023	26/01/2023
							F1905	Device Revision or Replacement	A090804	False Positive Result	26/01/2023	26/01/2023
12248	2023		47265	Dual-chamber implantable pacemaker, rate-responsive	CT141	Electrodes/Leads and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem		
							F1901	Additional Surgery	A24	No Apparent Device Problem		
					CT145	Cardiovascular devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem		
							F1901	Additional Surgery	A24	No Apparent Device Problem		
					CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem		
							F1901	Additional Surgery	A24	No Apparent Device Problem		
					CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem		
							F1901	Additional Surgery	A24	No Apparent Device Problem		
					CT233	Surgical	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem		
							F1901	Additional Surgery	A24	No Apparent Device Problem		
					CT2406	Implantable	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem		
							F1901	Additional Surgery	A24	No Apparent Device Problem		
					CT2407	Active implantable	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem		
							F1901	Additional Surgery	A24	No Apparent Device Problem		
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem		
							F1901	Additional Surgery	A24	No Apparent Device Problem		
					CT2727	Active implant devices/ancillaries	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem		
							F1901	Additional Surgery	A24	No Apparent Device Problem		
					CT2750	Implantable cardiac electrical stimulators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem		
							F1901	Additional Surgery	A24	No Apparent Device Problem		
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem		
							F1901	Additional Surgery	A24	No Apparent Device Problem		
					CT2985	Active	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem		
							F1901	Additional Surgery	A24	No Apparent Device Problem		
					CT3001	Implantable metal	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem		
							F1901	Additional Surgery	A24	No Apparent Device Problem		
					CT3002	Implantable synthetic polymer	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem		
							F1901	Additional Surgery	A24	No Apparent Device Problem		
					CT3006	Implantable inorganic material	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem		

							F1901	Additional Surgery	A24	No Apparent Device Problem		
					CT314	Electrically powered	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem		
							F1901	Additional Surgery	A24	No Apparent Device Problem		
					CT321	Long-term surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem		
							F1901	Additional Surgery	A24	No Apparent Device Problem		
					CT449	Generators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem		
							F1901	Additional Surgery	A24	No Apparent Device Problem		
					CT981	Single use	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem		
							F1901	Additional Surgery	A24	No Apparent Device Problem		
					CT983	Surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem		
							F1901	Additional Surgery	A24	No Apparent Device Problem		
12249	2023	All other reportable incidents	47265	Dual-chamber implantable pacemaker, rate-responsive	CT141	Electrodes/Leads and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
					CT145	Cardiovascular devices	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
					CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
					CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
					CT233	Surgical	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
					CT2406	Implantable	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
					CT2407	Active implantable	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
					CT2727	Active implant devices/ancillaries	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
					CT2750	Implantable cardiac electrical stimulators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
					CT2985	Active	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
					CT3001	Implantable metal	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
					CT3002	Implantable synthetic polymer	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
					CT3006	Implantable inorganic material	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
					CT314	Electrically powered	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
					CT321	Long-term surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
					CT449	Generators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
					CT981	Single use	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
					CT983	Surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem		

									A070504	Premature Discharge of Battery		
						F26	No Health Consequences or Impact		A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
12250	2023	All other reportable incidents	17882	Physiologic-monitoring defibrillation system	CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A07	Electrical /Electronic Property Problem		
							F27	Problem Identified During Non-Clinical Procedure	A07	Electrical /Electronic Property Problem		
12251	2023	All other reportable incidents	47265	Dual-chamber implantable pacemaker, rate-responsive	CT141	Electrodes/Leads and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	29/01/2023	29/01/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	29/01/2023	29/01/2023
					CT145	Cardiovascular devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	29/01/2023	29/01/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	29/01/2023	29/01/2023
					CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	29/01/2023	29/01/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	29/01/2023	29/01/2023
					CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	29/01/2023	29/01/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	29/01/2023	29/01/2023
					CT233	Surgical	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	29/01/2023	29/01/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	29/01/2023	29/01/2023
					CT2406	Implantable	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	29/01/2023	29/01/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	29/01/2023	29/01/2023
					CT2407	Active implantable	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	29/01/2023	29/01/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	29/01/2023	29/01/2023
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	29/01/2023	29/01/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	29/01/2023	29/01/2023
					CT2727	Active implant devices/ancillaries	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	29/01/2023	29/01/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	29/01/2023	29/01/2023
					CT2750	Implantable cardiac electrical stimulators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	29/01/2023	29/01/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	29/01/2023	29/01/2023
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	29/01/2023	29/01/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	29/01/2023	29/01/2023
					CT2985	Active	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	29/01/2023	29/01/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	29/01/2023	29/01/2023
					CT3001	Implantable metal	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	29/01/2023	29/01/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	29/01/2023	29/01/2023
					CT3002	Implantable synthetic polymer	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	29/01/2023	29/01/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	29/01/2023	29/01/2023
					CT3006	Implantable inorganic material	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	29/01/2023	29/01/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	29/01/2023	29/01/2023
					CT314	Electrically powered	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	29/01/2023	29/01/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	29/01/2023	29/01/2023
					CT321	Long-term surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	29/01/2023	29/01/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	29/01/2023	29/01/2023
					CT449	Generators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	29/01/2023	29/01/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	29/01/2023	29/01/2023
					CT981	Single use	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	29/01/2023	29/01/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	29/01/2023	29/01/2023
					CT983	Surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	29/01/2023	29/01/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	29/01/2023	29/01/2023
12252	2023	All other reportable incidents	35852	Single-chamber implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	20/02/2023	20/02/2023
							F1901	Additional Surgery	A070504	Premature Discharge of Battery	20/02/2023	20/02/2023
12253	2023		47265	Dual-chamber implantable pacemaker, rate-responsive	CT141	Electrodes/Leads and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	25/01/2023	25/01/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	25/01/2023	25/01/2023
					CT145	Cardiovascular devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	25/01/2023	25/01/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	25/01/2023	25/01/2023
					CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	25/01/2023	25/01/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	25/01/2023	25/01/2023
					CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	25/01/2023	25/01/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	25/01/2023	25/01/2023
					CT233	Surgical	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	25/01/2023	25/01/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	25/01/2023	25/01/2023
					CT2406	Implantable	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	25/01/2023	25/01/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	25/01/2023	25/01/2023
					CT2407	Active implantable	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	25/01/2023	25/01/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	25/01/2023	25/01/2023
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	25/01/2023	25/01/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	25/01/2023	25/01/2023
					CT2727	Active implant devices/ancillaries	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	25/01/2023	25/01/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	25/01/2023	25/01/2023
					CT2750	Implantable cardiac electrical stimulators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	25/01/2023	25/01/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	25/01/2023	25/01/2023
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	25/01/2023	25/01/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	25/01/2023	25/01/2023
					CT2985	Active	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	25/01/2023	25/01/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	25/01/2023	25/01/2023
					CT3001	Implantable metal	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	25/01/2023	25/01/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	25/01/2023	25/01/2023
					CT3002	Implantable synthetic polymer	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	25/01/2023	25/01/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	25/01/2023	25/01/2023
					CT3006	Implantable inorganic material	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	25/01/2023	25/01/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	25/01/2023	25/01/2023
					CT314	Electrically powered	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	25/01/2023	25/01/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	25/01/2023	25/01/2023
					CT321	Long-term surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	25/01/2023	25/01/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	25/01/2023	25/01/2023
					CT449	Generators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	25/01/2023	25/01/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	25/01/2023	25/01/2023
					CT981	Single use	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	25/01/2023	25/01/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	25/01/2023	25/01/2023
					CT983	Surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	25/01/2023	25/01/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	25/01/2023	25/01/2023
12254	2023		47265	Dual-chamber implantable pacemaker, rate-	CT141	Electrodes/Leads and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	27/01/2023	27/01/2023

				responsive			F1901	Additional Surgery	A24	No Apparent Device Problem	27/01/2023	27/01/2023
					CT145	Cardiovascular devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	27/01/2023	27/01/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	27/01/2023	27/01/2023
					CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	27/01/2023	27/01/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	27/01/2023	27/01/2023
					CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	27/01/2023	27/01/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	27/01/2023	27/01/2023
					CT233	Surgical	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	27/01/2023	27/01/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	27/01/2023	27/01/2023
					CT2406	Implantable	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	27/01/2023	27/01/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	27/01/2023	27/01/2023
					CT2407	Active implantable	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	27/01/2023	27/01/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	27/01/2023	27/01/2023
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	27/01/2023	27/01/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	27/01/2023	27/01/2023
					CT2727	Active implant devices/ancillaries	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	27/01/2023	27/01/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	27/01/2023	27/01/2023
					CT2750	Implantable cardiac electrical stimulators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	27/01/2023	27/01/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	27/01/2023	27/01/2023
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	27/01/2023	27/01/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	27/01/2023	27/01/2023
					CT2985	Active	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	27/01/2023	27/01/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	27/01/2023	27/01/2023
					CT3001	Implantable metal	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	27/01/2023	27/01/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	27/01/2023	27/01/2023
					CT3002	Implantable synthetic polymer	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	27/01/2023	27/01/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	27/01/2023	27/01/2023
					CT3006	Implantable inorganic material	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	27/01/2023	27/01/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	27/01/2023	27/01/2023
					CT314	Electrically powered	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	27/01/2023	27/01/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	27/01/2023	27/01/2023
					CT321	Long-term surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	27/01/2023	27/01/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	27/01/2023	27/01/2023
					CT449	Generators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	27/01/2023	27/01/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	27/01/2023	27/01/2023
					CT981	Single use	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	27/01/2023	27/01/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	27/01/2023	27/01/2023
					CT983	Surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	27/01/2023	27/01/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	27/01/2023	27/01/2023
12255	2023		47265	Dual-chamber implantable pacemaker, rate-responsive	CT141	Electrodes/Leads and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	05/02/2023	05/02/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	05/02/2023	05/02/2023
					CT145	Cardiovascular devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	05/02/2023	05/02/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	05/02/2023	05/02/2023
					CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	05/02/2023	05/02/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	05/02/2023	05/02/2023
					CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	05/02/2023	05/02/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	05/02/2023	05/02/2023
					CT233	Surgical	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	05/02/2023	05/02/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	05/02/2023	05/02/2023
					CT2406	Implantable	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	05/02/2023	05/02/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	05/02/2023	05/02/2023
					CT2407	Active implantable	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	05/02/2023	05/02/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	05/02/2023	05/02/2023
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	05/02/2023	05/02/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	05/02/2023	05/02/2023
					CT2727	Active implant devices/ancillaries	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	05/02/2023	05/02/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	05/02/2023	05/02/2023
					CT2750	Implantable cardiac electrical stimulators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	05/02/2023	05/02/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	05/02/2023	05/02/2023
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	05/02/2023	05/02/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	05/02/2023	05/02/2023
					CT2985	Active	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	05/02/2023	05/02/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	05/02/2023	05/02/2023
					CT3001	Implantable metal	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	05/02/2023	05/02/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	05/02/2023	05/02/2023
					CT3002	Implantable synthetic polymer	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	05/02/2023	05/02/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	05/02/2023	05/02/2023
					CT3006	Implantable inorganic material	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	05/02/2023	05/02/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	05/02/2023	05/02/2023
					CT314	Electrically powered	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	05/02/2023	05/02/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	05/02/2023	05/02/2023
					CT321	Long-term surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	05/02/2023	05/02/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	05/02/2023	05/02/2023
					CT449	Generators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	05/02/2023	05/02/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	05/02/2023	05/02/2023
					CT981	Single use	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	05/02/2023	05/02/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	05/02/2023	05/02/2023
					CT983	Surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	05/02/2023	05/02/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	05/02/2023	05/02/2023
12256	2023		47265	Dual-chamber implantable pacemaker, rate-responsive	CT141	Electrodes/Leads and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	19/01/2023	19/01/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	19/01/2023	19/01/2023
					CT145	Cardiovascular devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	19/01/2023	19/01/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	19/01/2023	19/01/2023
					CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	19/01/2023	19/01/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	19/01/2023	19/01/2023
					CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	19/01/2023	19/01/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	19/01/2023	19/01/2023
					CT233	Surgical	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	19/01/2023	19/01/2023

							F1901	Additional Surgery	A24	No Apparent Device Problem	19/01/2023	19/01/2023
					CT2406	Implantable	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	19/01/2023	19/01/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	19/01/2023	19/01/2023
					CT2407	Active implantable	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	19/01/2023	19/01/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	19/01/2023	19/01/2023
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	19/01/2023	19/01/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	19/01/2023	19/01/2023
					CT2727	Active implant devices/ancillaries	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	19/01/2023	19/01/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	19/01/2023	19/01/2023
					CT2750	Implantable cardiac electrical stimulators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	19/01/2023	19/01/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	19/01/2023	19/01/2023
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	19/01/2023	19/01/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	19/01/2023	19/01/2023
					CT2985	Active	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	19/01/2023	19/01/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	19/01/2023	19/01/2023
					CT3001	Implantable metal	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	19/01/2023	19/01/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	19/01/2023	19/01/2023
					CT3002	Implantable synthetic polymer	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	19/01/2023	19/01/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	19/01/2023	19/01/2023
					CT3006	Implantable inorganic material	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	19/01/2023	19/01/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	19/01/2023	19/01/2023
					CT314	Electrically powered	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	19/01/2023	19/01/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	19/01/2023	19/01/2023
					CT321	Long-term surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	19/01/2023	19/01/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	19/01/2023	19/01/2023
					CT449	Generators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	19/01/2023	19/01/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	19/01/2023	19/01/2023
					CT981	Single use	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	19/01/2023	19/01/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	19/01/2023	19/01/2023
					CT983	Surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	19/01/2023	19/01/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	19/01/2023	19/01/2023
12257	2023		47265	Dual-chamber implantable pacemaker, rate-responsive	CT141	Electrodes/Leads and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	06/02/2023	06/02/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	06/02/2023	06/02/2023
					CT145	Cardiovascular devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	06/02/2023	06/02/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	06/02/2023	06/02/2023
					CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	06/02/2023	06/02/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	06/02/2023	06/02/2023
					CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	06/02/2023	06/02/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	06/02/2023	06/02/2023
					CT233	Surgical	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	06/02/2023	06/02/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	06/02/2023	06/02/2023
					CT2406	Implantable	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	06/02/2023	06/02/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	06/02/2023	06/02/2023
					CT2407	Active implantable	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	06/02/2023	06/02/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	06/02/2023	06/02/2023
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	06/02/2023	06/02/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	06/02/2023	06/02/2023
					CT2727	Active implant devices/ancillaries	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	06/02/2023	06/02/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	06/02/2023	06/02/2023
					CT2750	Implantable cardiac electrical stimulators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	06/02/2023	06/02/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	06/02/2023	06/02/2023
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	06/02/2023	06/02/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	06/02/2023	06/02/2023
					CT2985	Active	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	06/02/2023	06/02/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	06/02/2023	06/02/2023
					CT3001	Implantable metal	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	06/02/2023	06/02/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	06/02/2023	06/02/2023
					CT3002	Implantable synthetic polymer	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	06/02/2023	06/02/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	06/02/2023	06/02/2023
					CT3006	Implantable inorganic material	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	06/02/2023	06/02/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	06/02/2023	06/02/2023
					CT314	Electrically powered	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	06/02/2023	06/02/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	06/02/2023	06/02/2023
					CT321	Long-term surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	06/02/2023	06/02/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	06/02/2023	06/02/2023
					CT449	Generators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	06/02/2023	06/02/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	06/02/2023	06/02/2023
					CT981	Single use	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	06/02/2023	06/02/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	06/02/2023	06/02/2023
					CT983	Surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	06/02/2023	06/02/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	06/02/2023	06/02/2023
12258	2023		47265	Dual-chamber implantable pacemaker, rate-responsive	CT141	Electrodes/Leads and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	27/01/2023	27/01/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	27/01/2023	27/01/2023
					CT145	Cardiovascular devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	27/01/2023	27/01/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	27/01/2023	27/01/2023
					CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	27/01/2023	27/01/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	27/01/2023	27/01/2023
					CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	27/01/2023	27/01/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	27/01/2023	27/01/2023
					CT233	Surgical	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	27/01/2023	27/01/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	27/01/2023	27/01/2023
					CT2406	Implantable	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	27/01/2023	27/01/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	27/01/2023	27/01/2023
					CT2407	Active implantable	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	27/01/2023	27/01/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	27/01/2023	27/01/2023
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	27/01/2023	27/01/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	27/01/2023	27/01/2023
					CT2727	Active implant devices/ancillaries	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	27/01/2023	27/01/2023

							F1901	Additional Surgery	A24	No Apparent Device Problem	27/01/2023	27/01/2023
					CT2750	Implantable cardiac electrical stimulators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	27/01/2023	27/01/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	27/01/2023	27/01/2023
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	27/01/2023	27/01/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	27/01/2023	27/01/2023
					CT2985	Active	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	27/01/2023	27/01/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	27/01/2023	27/01/2023
					CT3001	Implantable metal	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	27/01/2023	27/01/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	27/01/2023	27/01/2023
					CT3002	Implantable synthetic polymer	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	27/01/2023	27/01/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	27/01/2023	27/01/2023
					CT3006	Implantable inorganic material	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	27/01/2023	27/01/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	27/01/2023	27/01/2023
					CT314	Electrically powered	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	27/01/2023	27/01/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	27/01/2023	27/01/2023
					CT321	Long-term surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	27/01/2023	27/01/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	27/01/2023	27/01/2023
					CT449	Generators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	27/01/2023	27/01/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	27/01/2023	27/01/2023
					CT981	Single use	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	27/01/2023	27/01/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	27/01/2023	27/01/2023
					CT983	Surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	27/01/2023	27/01/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	27/01/2023	27/01/2023
12259	2023		47265	Dual-chamber implantable pacemaker, rate-responsive	CT141	Electrodes/Leads and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	06/02/2023	06/02/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	06/02/2023	06/02/2023
					CT145	Cardiovascular devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	06/02/2023	06/02/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	06/02/2023	06/02/2023
					CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	06/02/2023	06/02/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	06/02/2023	06/02/2023
					CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	06/02/2023	06/02/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	06/02/2023	06/02/2023
					CT233	Surgical	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	06/02/2023	06/02/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	06/02/2023	06/02/2023
					CT2406	Implantable	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	06/02/2023	06/02/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	06/02/2023	06/02/2023
					CT2407	Active implantable	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	06/02/2023	06/02/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	06/02/2023	06/02/2023
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	06/02/2023	06/02/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	06/02/2023	06/02/2023
					CT2727	Active implant devices/ancillaries	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	06/02/2023	06/02/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	06/02/2023	06/02/2023
					CT2750	Implantable cardiac electrical stimulators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	06/02/2023	06/02/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	06/02/2023	06/02/2023
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	06/02/2023	06/02/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	06/02/2023	06/02/2023
					CT2985	Active	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	06/02/2023	06/02/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	06/02/2023	06/02/2023
					CT3001	Implantable metal	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	06/02/2023	06/02/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	06/02/2023	06/02/2023
					CT3002	Implantable synthetic polymer	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	06/02/2023	06/02/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	06/02/2023	06/02/2023
					CT3006	Implantable inorganic material	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	06/02/2023	06/02/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	06/02/2023	06/02/2023
					CT314	Electrically powered	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	06/02/2023	06/02/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	06/02/2023	06/02/2023
					CT321	Long-term surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	06/02/2023	06/02/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	06/02/2023	06/02/2023
					CT449	Generators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	06/02/2023	06/02/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	06/02/2023	06/02/2023
					CT981	Single use	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	06/02/2023	06/02/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	06/02/2023	06/02/2023
					CT983	Surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	06/02/2023	06/02/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	06/02/2023	06/02/2023
12260	2023	All other reportable incidents	63188	Left-ventricular wireless pacing system sensor/transmitter	CT141	Electrodes/Leads and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F24	Insufficient Health Impact Information	A070504	Premature Discharge of Battery		
					CT145	Cardiovascular devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F24	Insufficient Health Impact Information	A070504	Premature Discharge of Battery		
					CT1577	Wireless monitoring system transmitters	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F24	Insufficient Health Impact Information	A070504	Premature Discharge of Battery		
					CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F24	Insufficient Health Impact Information	A070504	Premature Discharge of Battery		
					CT2317	Transmitters/Receivers/Transceivers	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F24	Insufficient Health Impact Information	A070504	Premature Discharge of Battery		
					CT233	Surgical	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F24	Insufficient Health Impact Information	A070504	Premature Discharge of Battery		
					CT2406	Implantable	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F24	Insufficient Health Impact Information	A070504	Premature Discharge of Battery		
					CT2407	Active implantable	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F24	Insufficient Health Impact Information	A070504	Premature Discharge of Battery		
					CT255	Transducers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F24	Insufficient Health Impact Information	A070504	Premature Discharge of Battery		
					CT2726	Physiological monitoring/measuring devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F24	Insufficient Health Impact Information	A070504	Premature Discharge of Battery		
					CT2727	Active implant devices/ancillaries	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F24	Insufficient Health Impact Information	A070504	Premature Discharge of Battery		
					CT2750	Implantable cardiac electrical stimulators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F24	Insufficient Health Impact Information	A070504	Premature Discharge of Battery		
					CT2885	Physiological monitoring/measuring system	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		

						transmitters/receivers/transceivers	F24	Insufficient Health Impact Information	A070504	Premature Discharge of Battery		
					CT2985	Active	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F24	Insufficient Health Impact Information	A070504	Premature Discharge of Battery		
					CT3001	Implantable metal	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F24	Insufficient Health Impact Information	A070504	Premature Discharge of Battery		
					CT3002	Implantable synthetic polymer	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F24	Insufficient Health Impact Information	A070504	Premature Discharge of Battery		
					CT3006	Implantable inorganic material	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F24	Insufficient Health Impact Information	A070504	Premature Discharge of Battery		
					CT314	Electrically powered	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F24	Insufficient Health Impact Information	A070504	Premature Discharge of Battery		
					CT321	Long-term surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F24	Insufficient Health Impact Information	A070504	Premature Discharge of Battery		
					CT384	Sensors and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F24	Insufficient Health Impact Information	A070504	Premature Discharge of Battery		
					CT449	Generators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F24	Insufficient Health Impact Information	A070504	Premature Discharge of Battery		
					CT981	Single use	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F24	Insufficient Health Impact Information	A070504	Premature Discharge of Battery		
					CT983	Surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F24	Insufficient Health Impact Information	A070504	Premature Discharge of Battery		
					CT994	Sensors	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F24	Insufficient Health Impact Information	A070504	Premature Discharge of Battery		
12261	2023	All other reportable incidents	35223	Endocardial/interventricular septal pacing lead	CT1684	Cardiac pacemakers and associated devices	E0604	Cardiac Perforation	A24	No Apparent Device Problem		
							E0731	Pleural Effusion	A24	No Apparent Device Problem		
							E0734	Pneumothorax	A24	No Apparent Device Problem		
							F08	Hospitalization or Prolonged Hospitalization	A24	No Apparent Device Problem		
							F1203	Life Threatening Illness or Injury	A24	No Apparent Device Problem		
							F1904	Device Repositioning	A24	No Apparent Device Problem		
12262	2023	All other reportable incidents	35223	Endocardial/interventricular septal pacing lead	CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A07	Electrical /Electronic Property Problem		
									A070101	Failure to Capture		
							F26	No Health Consequences or Impact	A07	Electrical /Electronic Property Problem		
									A070101	Failure to Capture		
12263	2023	All other reportable incidents	35223	Endocardial/interventricular septal pacing lead	CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A05	Mechanical Problem		
									A070909	Oversensing		
									A090801	Signal Artifact/Noise		
							F26	No Health Consequences or Impact	A05	Mechanical Problem		
									A070909	Oversensing		
									A090801	Signal Artifact/Noise		
12264	2023	All other reportable incidents	35223	Endocardial/interventricular septal pacing lead	CT1684	Cardiac pacemakers and associated devices	E011903	Syncope/Fainting	A070909	Oversensing		
							E060106	Heart Block	A070909	Oversensing		
							F08	Hospitalization or Prolonged Hospitalization	A070909	Oversensing		
							F1203	Life Threatening Illness or Injury	A070909	Oversensing		
12265	2023	All other reportable incidents	47265	Dual-chamber implantable pacemaker, rate-responsive	CT141	Electrodes/Leads and associated devices	E060106	Heart Block	A0712	Pacing Problem		
							F12	Serious Injury/ Illness/ Impairment	A0712	Pacing Problem		
							F23	Unexpected Medical Intervention	A0712	Pacing Problem		
					CT145	Cardiovascular devices	E060106	Heart Block	A0712	Pacing Problem		
							F12	Serious Injury/ Illness/ Impairment	A0712	Pacing Problem		
							F23	Unexpected Medical Intervention	A0712	Pacing Problem		
					CT1684	Cardiac pacemakers and associated devices	E060106	Heart Block	A0712	Pacing Problem		
							F12	Serious Injury/ Illness/ Impairment	A0712	Pacing Problem		
							F23	Unexpected Medical Intervention	A0712	Pacing Problem		
					CT1685	Cardiac pacemakers	E060106	Heart Block	A0712	Pacing Problem		
							F12	Serious Injury/ Illness/ Impairment	A0712	Pacing Problem		
							F23	Unexpected Medical Intervention	A0712	Pacing Problem		
					CT233	Surgical	E060106	Heart Block	A0712	Pacing Problem		
							F12	Serious Injury/ Illness/ Impairment	A0712	Pacing Problem		
							F23	Unexpected Medical Intervention	A0712	Pacing Problem		
					CT2406	Implantable	E060106	Heart Block	A0712	Pacing Problem		
							F12	Serious Injury/ Illness/ Impairment	A0712	Pacing Problem		
							F23	Unexpected Medical Intervention	A0712	Pacing Problem		
					CT2407	Active implantable	E060106	Heart Block	A0712	Pacing Problem		
							F12	Serious Injury/ Illness/ Impairment	A0712	Pacing Problem		
							F23	Unexpected Medical Intervention	A0712	Pacing Problem		
					CT257	Implantable cardiac pacemakers	E060106	Heart Block	A0712	Pacing Problem		
							F12	Serious Injury/ Illness/ Impairment	A0712	Pacing Problem		
							F23	Unexpected Medical Intervention	A0712	Pacing Problem		
					CT2727	Active implant devices/ancillaries	E060106	Heart Block	A0712	Pacing Problem		
							F12	Serious Injury/ Illness/ Impairment	A0712	Pacing Problem		
							F23	Unexpected Medical Intervention	A0712	Pacing Problem		
					CT2750	Implantable cardiac electrical stimulators and associated devices	E060106	Heart Block	A0712	Pacing Problem		
							F12	Serious Injury/ Illness/ Impairment	A0712	Pacing Problem		
							F23	Unexpected Medical Intervention	A0712	Pacing Problem		
					CT2866	Implantable cardiac electrical stimulators	E060106	Heart Block	A0712	Pacing Problem		
							F12	Serious Injury/ Illness/ Impairment	A0712	Pacing Problem		
							F23	Unexpected Medical Intervention	A0712	Pacing Problem		
					CT2985	Active	E060106	Heart Block	A0712	Pacing Problem		
							F12	Serious Injury/ Illness/ Impairment	A0712	Pacing Problem		
							F23	Unexpected Medical Intervention	A0712	Pacing Problem		
					CT3001	Implantable metal	E060106	Heart Block	A0712	Pacing Problem		
							F12	Serious Injury/ Illness/ Impairment	A0712	Pacing Problem		
							F23	Unexpected Medical Intervention	A0712	Pacing Problem		
					CT3002	Implantable synthetic polymer	E060106	Heart Block	A0712	Pacing Problem		
							F12	Serious Injury/ Illness/ Impairment	A0712	Pacing Problem		
							F23	Unexpected Medical Intervention	A0712	Pacing Problem		
					CT3006	Implantable inorganic material	E060106	Heart Block	A0712	Pacing Problem		
							F12	Serious Injury/ Illness/ Impairment	A0712	Pacing Problem		
							F23	Unexpected Medical Intervention	A0712	Pacing Problem		

					CT314	Electrically powered	E060106	Heart Block	A0712	Pacing Problem		
							F12	Serious Injury/ Illness/ Impairment	A0712	Pacing Problem		
							F23	Unexpected Medical Intervention	A0712	Pacing Problem		
					CT321	Long-term surgical invasive	E060106	Heart Block	A0712	Pacing Problem		
							F12	Serious Injury/ Illness/ Impairment	A0712	Pacing Problem		
							F23	Unexpected Medical Intervention	A0712	Pacing Problem		
					CT449	Generators and associated devices	E060106	Heart Block	A0712	Pacing Problem		
							F12	Serious Injury/ Illness/ Impairment	A0712	Pacing Problem		
							F23	Unexpected Medical Intervention	A0712	Pacing Problem		
					CT981	Single use	E060106	Heart Block	A0712	Pacing Problem		
							F12	Serious Injury/ Illness/ Impairment	A0712	Pacing Problem		
							F23	Unexpected Medical Intervention	A0712	Pacing Problem		
					CT983	Surgical invasive	E060106	Heart Block	A0712	Pacing Problem		
							F12	Serious Injury/ Illness/ Impairment	A0712	Pacing Problem		
							F23	Unexpected Medical Intervention	A0712	Pacing Problem		
12266	2023	All other reportable incidents	47265	Dual-chamber implantable pacemaker, rate-responsive	CT141	Electrodes/Leads and associated devices	E060106	Heart Block	A0712	Pacing Problem		
							F08	Hospitalization or Prolonged Hospitalization	A0712	Pacing Problem		
							F12	Serious Injury/ Illness/ Impairment	A0712	Pacing Problem		
							F19	Surgical Intervention	A0712	Pacing Problem		
							F23	Unexpected Medical Intervention	A0712	Pacing Problem		
					CT145	Cardiovascular devices	E060106	Heart Block	A0712	Pacing Problem		
							F08	Hospitalization or Prolonged Hospitalization	A0712	Pacing Problem		
							F12	Serious Injury/ Illness/ Impairment	A0712	Pacing Problem		
							F19	Surgical Intervention	A0712	Pacing Problem		
							F23	Unexpected Medical Intervention	A0712	Pacing Problem		
					CT1684	Cardiac pacemakers and associated devices	E060106	Heart Block	A0712	Pacing Problem		
							F08	Hospitalization or Prolonged Hospitalization	A0712	Pacing Problem		
							F12	Serious Injury/ Illness/ Impairment	A0712	Pacing Problem		
							F19	Surgical Intervention	A0712	Pacing Problem		
							F23	Unexpected Medical Intervention	A0712	Pacing Problem		
					CT1685	Cardiac pacemakers	E060106	Heart Block	A0712	Pacing Problem		
							F08	Hospitalization or Prolonged Hospitalization	A0712	Pacing Problem		
							F12	Serious Injury/ Illness/ Impairment	A0712	Pacing Problem		
							F19	Surgical Intervention	A0712	Pacing Problem		
							F23	Unexpected Medical Intervention	A0712	Pacing Problem		
					CT233	Surgical	E060106	Heart Block	A0712	Pacing Problem		
							F08	Hospitalization or Prolonged Hospitalization	A0712	Pacing Problem		
							F12	Serious Injury/ Illness/ Impairment	A0712	Pacing Problem		
							F19	Surgical Intervention	A0712	Pacing Problem		
							F23	Unexpected Medical Intervention	A0712	Pacing Problem		
					CT2406	Implantable	E060106	Heart Block	A0712	Pacing Problem		
							F08	Hospitalization or Prolonged Hospitalization	A0712	Pacing Problem		
							F12	Serious Injury/ Illness/ Impairment	A0712	Pacing Problem		
							F19	Surgical Intervention	A0712	Pacing Problem		
							F23	Unexpected Medical Intervention	A0712	Pacing Problem		
					CT2407	Active implantable	E060106	Heart Block	A0712	Pacing Problem		
							F08	Hospitalization or Prolonged Hospitalization	A0712	Pacing Problem		
							F12	Serious Injury/ Illness/ Impairment	A0712	Pacing Problem		
							F19	Surgical Intervention	A0712	Pacing Problem		
							F23	Unexpected Medical Intervention	A0712	Pacing Problem		
					CT257	Implantable cardiac pacemakers	E060106	Heart Block	A0712	Pacing Problem		
							F08	Hospitalization or Prolonged Hospitalization	A0712	Pacing Problem		
							F12	Serious Injury/ Illness/ Impairment	A0712	Pacing Problem		
							F19	Surgical Intervention	A0712	Pacing Problem		
							F23	Unexpected Medical Intervention	A0712	Pacing Problem		
					CT2727	Active implant devices/ancillaries	E060106	Heart Block	A0712	Pacing Problem		
							F08	Hospitalization or Prolonged Hospitalization	A0712	Pacing Problem		
							F12	Serious Injury/ Illness/ Impairment	A0712	Pacing Problem		
							F19	Surgical Intervention	A0712	Pacing Problem		
							F23	Unexpected Medical Intervention	A0712	Pacing Problem		
					CT2750	Implantable cardiac electrical stimulators and associated devices	E060106	Heart Block	A0712	Pacing Problem		
							F08	Hospitalization or Prolonged Hospitalization	A0712	Pacing Problem		
							F12	Serious Injury/ Illness/ Impairment	A0712	Pacing Problem		
							F19	Surgical Intervention	A0712	Pacing Problem		
							F23	Unexpected Medical Intervention	A0712	Pacing Problem		
					CT2866	Implantable cardiac electrical stimulators	E060106	Heart Block	A0712	Pacing Problem		
							F08	Hospitalization or Prolonged Hospitalization	A0712	Pacing Problem		
							F12	Serious Injury/ Illness/ Impairment	A0712	Pacing Problem		
							F19	Surgical Intervention	A0712	Pacing Problem		
							F23	Unexpected Medical Intervention	A0712	Pacing Problem		
					CT2985	Active	E060106	Heart Block	A0712	Pacing Problem		
							F08	Hospitalization or Prolonged Hospitalization	A0712	Pacing Problem		
							F12	Serious Injury/ Illness/ Impairment	A0712	Pacing Problem		
							F19	Surgical Intervention	A0712	Pacing Problem		
							F23	Unexpected Medical Intervention	A0712	Pacing Problem		
					CT3001	Implantable metal	E060106	Heart Block	A0712	Pacing Problem		
							F08	Hospitalization or Prolonged Hospitalization	A0712	Pacing Problem		
							F12	Serious Injury/ Illness/ Impairment	A0712	Pacing Problem		
							F19	Surgical Intervention	A0712	Pacing Problem		
							F23	Unexpected Medical Intervention	A0712	Pacing Problem		
					CT3002	Implantable synthetic polymer	E060106	Heart Block	A0712	Pacing Problem		
							F08	Hospitalization or Prolonged Hospitalization	A0712	Pacing Problem		
							F12	Serious Injury/ Illness/ Impairment	A0712	Pacing Problem		
							F19	Surgical Intervention	A0712	Pacing Problem		
							F23	Unexpected Medical Intervention	A0712	Pacing Problem		
					CT3006	Implantable inorganic material	E060106	Heart Block	A0712	Pacing Problem		
							F08	Hospitalization or Prolonged Hospitalization	A0712	Pacing Problem		
							F12	Serious Injury/ Illness/ Impairment	A0712	Pacing Problem		

						F19	Surgical Intervention	A0712	Pacing Problem			
						F23	Unexpected Medical Intervention	A0712	Pacing Problem			
					CT314	Electrically powered	E060106	Heart Block	A0712	Pacing Problem		
							F08	Hospitalization or Prolonged Hospitalization	A0712	Pacing Problem		
							F12	Serious Injury/ Illness/ Impairment	A0712	Pacing Problem		
							F19	Surgical Intervention	A0712	Pacing Problem		
							F23	Unexpected Medical Intervention	A0712	Pacing Problem		
					CT321	Long-term surgical invasive	E060106	Heart Block	A0712	Pacing Problem		
							F08	Hospitalization or Prolonged Hospitalization	A0712	Pacing Problem		
							F12	Serious Injury/ Illness/ Impairment	A0712	Pacing Problem		
							F19	Surgical Intervention	A0712	Pacing Problem		
							F23	Unexpected Medical Intervention	A0712	Pacing Problem		
					CT449	Generators and associated devices	E060106	Heart Block	A0712	Pacing Problem		
							F08	Hospitalization or Prolonged Hospitalization	A0712	Pacing Problem		
							F12	Serious Injury/ Illness/ Impairment	A0712	Pacing Problem		
							F19	Surgical Intervention	A0712	Pacing Problem		
							F23	Unexpected Medical Intervention	A0712	Pacing Problem		
					CT981	Single use	E060106	Heart Block	A0712	Pacing Problem		
							F08	Hospitalization or Prolonged Hospitalization	A0712	Pacing Problem		
							F12	Serious Injury/ Illness/ Impairment	A0712	Pacing Problem		
							F19	Surgical Intervention	A0712	Pacing Problem		
							F23	Unexpected Medical Intervention	A0712	Pacing Problem		
					CT983	Surgical invasive	E060106	Heart Block	A0712	Pacing Problem		
							F08	Hospitalization or Prolonged Hospitalization	A0712	Pacing Problem		
							F12	Serious Injury/ Illness/ Impairment	A0712	Pacing Problem		
							F19	Surgical Intervention	A0712	Pacing Problem		
							F23	Unexpected Medical Intervention	A0712	Pacing Problem		
12267	2023	All other reportable incidents	47265	Dual-chamber implantable pacemaker, rate-responsive	CT141	Electrodes/Leads and associated devices	E060106	Heart Block	A0712	Pacing Problem		
							F12	Serious Injury/ Illness/ Impairment	A0712	Pacing Problem		
							F23	Unexpected Medical Intervention	A0712	Pacing Problem		
					CT145	Cardiovascular devices	E060106	Heart Block	A0712	Pacing Problem		
							F12	Serious Injury/ Illness/ Impairment	A0712	Pacing Problem		
							F23	Unexpected Medical Intervention	A0712	Pacing Problem		
					CT1684	Cardiac pacemakers and associated devices	E060106	Heart Block	A0712	Pacing Problem		
							F12	Serious Injury/ Illness/ Impairment	A0712	Pacing Problem		
							F23	Unexpected Medical Intervention	A0712	Pacing Problem		
					CT1685	Cardiac pacemakers	E060106	Heart Block	A0712	Pacing Problem		
							F12	Serious Injury/ Illness/ Impairment	A0712	Pacing Problem		
							F23	Unexpected Medical Intervention	A0712	Pacing Problem		
					CT233	Surgical	E060106	Heart Block	A0712	Pacing Problem		
							F12	Serious Injury/ Illness/ Impairment	A0712	Pacing Problem		
							F23	Unexpected Medical Intervention	A0712	Pacing Problem		
					CT2406	Implantable	E060106	Heart Block	A0712	Pacing Problem		
							F12	Serious Injury/ Illness/ Impairment	A0712	Pacing Problem		
							F23	Unexpected Medical Intervention	A0712	Pacing Problem		
					CT2407	Active implantable	E060106	Heart Block	A0712	Pacing Problem		
							F12	Serious Injury/ Illness/ Impairment	A0712	Pacing Problem		
							F23	Unexpected Medical Intervention	A0712	Pacing Problem		
					CT257	Implantable cardiac pacemakers	E060106	Heart Block	A0712	Pacing Problem		
							F12	Serious Injury/ Illness/ Impairment	A0712	Pacing Problem		
							F23	Unexpected Medical Intervention	A0712	Pacing Problem		
					CT2727	Active implant devices/ancillaries	E060106	Heart Block	A0712	Pacing Problem		
							F12	Serious Injury/ Illness/ Impairment	A0712	Pacing Problem		
							F23	Unexpected Medical Intervention	A0712	Pacing Problem		
					CT2750	Implantable cardiac electrical stimulators and associated devices	E060106	Heart Block	A0712	Pacing Problem		
							F12	Serious Injury/ Illness/ Impairment	A0712	Pacing Problem		
							F23	Unexpected Medical Intervention	A0712	Pacing Problem		
					CT2866	Implantable cardiac electrical stimulators	E060106	Heart Block	A0712	Pacing Problem		
							F12	Serious Injury/ Illness/ Impairment	A0712	Pacing Problem		
							F23	Unexpected Medical Intervention	A0712	Pacing Problem		
					CT2985	Active	E060106	Heart Block	A0712	Pacing Problem		
							F12	Serious Injury/ Illness/ Impairment	A0712	Pacing Problem		
							F23	Unexpected Medical Intervention	A0712	Pacing Problem		
					CT3001	Implantable metal	E060106	Heart Block	A0712	Pacing Problem		
							F12	Serious Injury/ Illness/ Impairment	A0712	Pacing Problem		
							F23	Unexpected Medical Intervention	A0712	Pacing Problem		
					CT3002	Implantable synthetic polymer	E060106	Heart Block	A0712	Pacing Problem		
							F12	Serious Injury/ Illness/ Impairment	A0712	Pacing Problem		
							F23	Unexpected Medical Intervention	A0712	Pacing Problem		
					CT3006	Implantable inorganic material	E060106	Heart Block	A0712	Pacing Problem		
							F12	Serious Injury/ Illness/ Impairment	A0712	Pacing Problem		
							F23	Unexpected Medical Intervention	A0712	Pacing Problem		
					CT314	Electrically powered	E060106	Heart Block	A0712	Pacing Problem		
							F12	Serious Injury/ Illness/ Impairment	A0712	Pacing Problem		
							F23	Unexpected Medical Intervention	A0712	Pacing Problem		
					CT321	Long-term surgical invasive	E060106	Heart Block	A0712	Pacing Problem		
							F12	Serious Injury/ Illness/ Impairment	A0712	Pacing Problem		
							F23	Unexpected Medical Intervention	A0712	Pacing Problem		
					CT449	Generators and associated devices	E060106	Heart Block	A0712	Pacing Problem		
							F12	Serious Injury/ Illness/ Impairment	A0712	Pacing Problem		
							F23	Unexpected Medical Intervention	A0712	Pacing Problem		
					CT981	Single use	E060106	Heart Block	A0712	Pacing Problem		
							F12	Serious Injury/ Illness/ Impairment	A0712	Pacing Problem		
							F23	Unexpected Medical Intervention	A0712	Pacing Problem		
					CT983	Surgical invasive	E060106	Heart Block	A0712	Pacing Problem		
							F12	Serious Injury/ Illness/ Impairment	A0712	Pacing Problem		
							F23	Unexpected Medical Intervention	A0712	Pacing Problem		
12268	2023	All other reportable incidents	37265	Dual-chamber implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A070909	Oversensing		

							F26	No Health Consequences or Impact	A070909	Oversensing			
12269	2023		47265	Dual-chamber implantable pacemaker, rate-responsive	CT141	Electrodes/Leads and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem			
							F1901	Additional Surgery	A24	No Apparent Device Problem			
					CT145	Cardiovascular devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem			
							F1901	Additional Surgery	A24	No Apparent Device Problem			
					CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem			
							F1901	Additional Surgery	A24	No Apparent Device Problem			
					CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem			
							F1901	Additional Surgery	A24	No Apparent Device Problem			
					CT233	Surgical	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem			
							F1901	Additional Surgery	A24	No Apparent Device Problem			
					CT2406	Implantable	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem			
							F1901	Additional Surgery	A24	No Apparent Device Problem			
					CT2407	Active implantable	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem			
							F1901	Additional Surgery	A24	No Apparent Device Problem			
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem			
							F1901	Additional Surgery	A24	No Apparent Device Problem			
					CT2727	Active implant devices/ancillaries	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem			
							F1901	Additional Surgery	A24	No Apparent Device Problem			
					CT2750	Implantable cardiac electrical stimulators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem			
							F1901	Additional Surgery	A24	No Apparent Device Problem			
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem			
							F1901	Additional Surgery	A24	No Apparent Device Problem			
					CT2985	Active	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem			
							F1901	Additional Surgery	A24	No Apparent Device Problem			
					CT3001	Implantable metal	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem			
							F1901	Additional Surgery	A24	No Apparent Device Problem			
					CT3002	Implantable synthetic polymer	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem			
							F1901	Additional Surgery	A24	No Apparent Device Problem			
					CT3006	Implantable inorganic material	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem			
							F1901	Additional Surgery	A24	No Apparent Device Problem			
					CT314	Electrically powered	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem			
							F1901	Additional Surgery	A24	No Apparent Device Problem			
					CT321	Long-term surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem			
							F1901	Additional Surgery	A24	No Apparent Device Problem			
					CT449	Generators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem			
							F1901	Additional Surgery	A24	No Apparent Device Problem			
					CT981	Single use	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem			
							F1901	Additional Surgery	A24	No Apparent Device Problem			
					CT983	Surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem			
							F1901	Additional Surgery	A24	No Apparent Device Problem			
12270	2023		47265	Dual-chamber implantable pacemaker, rate-responsive	CT141	Electrodes/Leads and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	26/01/2023	26/01/2023	
							F1901	Additional Surgery	A24	No Apparent Device Problem	26/01/2023	26/01/2023	
					CT145	Cardiovascular devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	26/01/2023	26/01/2023	
							F1901	Additional Surgery	A24	No Apparent Device Problem	26/01/2023	26/01/2023	
					CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	26/01/2023	26/01/2023	
							F1901	Additional Surgery	A24	No Apparent Device Problem	26/01/2023	26/01/2023	
					CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	26/01/2023	26/01/2023	
							F1901	Additional Surgery	A24	No Apparent Device Problem	26/01/2023	26/01/2023	
					CT233	Surgical	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	26/01/2023	26/01/2023	
							F1901	Additional Surgery	A24	No Apparent Device Problem	26/01/2023	26/01/2023	
					CT2406	Implantable	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	26/01/2023	26/01/2023	
							F1901	Additional Surgery	A24	No Apparent Device Problem	26/01/2023	26/01/2023	
					CT2407	Active implantable	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	26/01/2023	26/01/2023	
							F1901	Additional Surgery	A24	No Apparent Device Problem	26/01/2023	26/01/2023	
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	26/01/2023	26/01/2023	
							F1901	Additional Surgery	A24	No Apparent Device Problem	26/01/2023	26/01/2023	
					CT2727	Active implant devices/ancillaries	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	26/01/2023	26/01/2023	
							F1901	Additional Surgery	A24	No Apparent Device Problem	26/01/2023	26/01/2023	
					CT2750	Implantable cardiac electrical stimulators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	26/01/2023	26/01/2023	
							F1901	Additional Surgery	A24	No Apparent Device Problem	26/01/2023	26/01/2023	
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	26/01/2023	26/01/2023	
							F1901	Additional Surgery	A24	No Apparent Device Problem	26/01/2023	26/01/2023	
					CT2985	Active	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	26/01/2023	26/01/2023	
							F1901	Additional Surgery	A24	No Apparent Device Problem	26/01/2023	26/01/2023	
					CT3001	Implantable metal	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	26/01/2023	26/01/2023	
							F1901	Additional Surgery	A24	No Apparent Device Problem	26/01/2023	26/01/2023	
					CT3002	Implantable synthetic polymer	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	26/01/2023	26/01/2023	
							F1901	Additional Surgery	A24	No Apparent Device Problem	26/01/2023	26/01/2023	
					CT3006	Implantable inorganic material	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	26/01/2023	26/01/2023	
							F1901	Additional Surgery	A24	No Apparent Device Problem	26/01/2023	26/01/2023	
					CT314	Electrically powered	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	26/01/2023	26/01/2023	
							F1901	Additional Surgery	A24	No Apparent Device Problem	26/01/2023	26/01/2023	
					CT321	Long-term surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	26/01/2023	26/01/2023	
							F1901	Additional Surgery	A24	No Apparent Device Problem	26/01/2023	26/01/2023	
					CT449	Generators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	26/01/2023	26/01/2023	
							F1901	Additional Surgery	A24	No Apparent Device Problem	26/01/2023	26/01/2023	
					CT981	Single use	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	26/01/2023	26/01/2023	
							F1901	Additional Surgery	A24	No Apparent Device Problem	26/01/2023	26/01/2023	
					CT983	Surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	26/01/2023	26/01/2023	
							F1901	Additional Surgery	A24	No Apparent Device Problem	26/01/2023	26/01/2023	
12271	2023		47265	Dual-chamber implantable pacemaker, rate-responsive	CT141	Electrodes/Leads and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	07/02/2023	07/02/2023	
							F1901	Additional Surgery	A24	No Apparent Device Problem	07/02/2023	07/02/2023	
					CT145	Cardiovascular devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	07/02/2023	07/02/2023	
							F1901	Additional Surgery	A24	No Apparent Device Problem	07/02/2023	07/02/2023	
					CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	07/02/2023	07/02/2023	
							F1901	Additional Surgery	A24	No Apparent Device Problem	07/02/2023	07/02/2023	
					CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	07/02/2023	07/02/2023	

							F1901	Additional Surgery	A24	No Apparent Device Problem	07/02/2023	07/02/2023
					CT233	Surgical	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	07/02/2023	07/02/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	07/02/2023	07/02/2023
					CT2406	Implantable	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	07/02/2023	07/02/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	07/02/2023	07/02/2023
					CT2407	Active implantable	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	07/02/2023	07/02/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	07/02/2023	07/02/2023
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	07/02/2023	07/02/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	07/02/2023	07/02/2023
					CT2727	Active implant devices/ancillaries	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	07/02/2023	07/02/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	07/02/2023	07/02/2023
					CT2750	Implantable cardiac electrical stimulators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	07/02/2023	07/02/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	07/02/2023	07/02/2023
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	07/02/2023	07/02/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	07/02/2023	07/02/2023
					CT2985	Active	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	07/02/2023	07/02/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	07/02/2023	07/02/2023
					CT3001	Implantable metal	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	07/02/2023	07/02/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	07/02/2023	07/02/2023
					CT3002	Implantable synthetic polymer	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	07/02/2023	07/02/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	07/02/2023	07/02/2023
					CT3006	Implantable inorganic material	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	07/02/2023	07/02/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	07/02/2023	07/02/2023
					CT314	Electrically powered	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	07/02/2023	07/02/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	07/02/2023	07/02/2023
					CT321	Long-term surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	07/02/2023	07/02/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	07/02/2023	07/02/2023
					CT449	Generators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	07/02/2023	07/02/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	07/02/2023	07/02/2023
					CT981	Single use	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	07/02/2023	07/02/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	07/02/2023	07/02/2023
					CT983	Surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	07/02/2023	07/02/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	07/02/2023	07/02/2023
12272	2023		47265	Dual-chamber implantable pacemaker, rate-responsive	CT141	Electrodes/Leads and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	31/01/2023	31/01/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	31/01/2023	31/01/2023
					CT145	Cardiovascular devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	31/01/2023	31/01/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	31/01/2023	31/01/2023
					CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	31/01/2023	31/01/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	31/01/2023	31/01/2023
					CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	31/01/2023	31/01/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	31/01/2023	31/01/2023
					CT233	Surgical	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	31/01/2023	31/01/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	31/01/2023	31/01/2023
					CT2406	Implantable	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	31/01/2023	31/01/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	31/01/2023	31/01/2023
					CT2407	Active implantable	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	31/01/2023	31/01/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	31/01/2023	31/01/2023
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	31/01/2023	31/01/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	31/01/2023	31/01/2023
					CT2727	Active implant devices/ancillaries	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	31/01/2023	31/01/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	31/01/2023	31/01/2023
					CT2750	Implantable cardiac electrical stimulators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	31/01/2023	31/01/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	31/01/2023	31/01/2023
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	31/01/2023	31/01/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	31/01/2023	31/01/2023
					CT2985	Active	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	31/01/2023	31/01/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	31/01/2023	31/01/2023
					CT3001	Implantable metal	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	31/01/2023	31/01/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	31/01/2023	31/01/2023
					CT3002	Implantable synthetic polymer	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	31/01/2023	31/01/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	31/01/2023	31/01/2023
					CT3006	Implantable inorganic material	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	31/01/2023	31/01/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	31/01/2023	31/01/2023
					CT314	Electrically powered	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	31/01/2023	31/01/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	31/01/2023	31/01/2023
					CT321	Long-term surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	31/01/2023	31/01/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	31/01/2023	31/01/2023
					CT449	Generators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	31/01/2023	31/01/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	31/01/2023	31/01/2023
					CT981	Single use	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	31/01/2023	31/01/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	31/01/2023	31/01/2023
					CT983	Surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	31/01/2023	31/01/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	31/01/2023	31/01/2023
12273	2023	All other reportable incidents	17882	Physiologic-monitoring defibrillation system	CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070908	Failure to Sense		
							F27	Problem Identified During Non-Clinical Procedure	A070908	Failure to Sense		
12274	2023	All other reportable incidents	37265	Dual-chamber implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A072201	High Impedance	16/01/2023	16/01/2023
							F26	No Health Consequences or Impact	A072201	High Impedance	16/01/2023	16/01/2023
12275	2023	All other reportable incidents	47265	Dual-chamber implantable pacemaker, rate-responsive	CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A0908	Incorrect, Inadequate or Imprecise Result or Readings	07/02/2023	07/02/2023
							F08	Hospitalization or Prolonged Hospitalization	A0908	Incorrect, Inadequate or Imprecise Result or Readings	07/02/2023	07/02/2023
							F1905	Device Revision or Replacement	A0908	Incorrect, Inadequate or Imprecise Result or Readings	07/02/2023	07/02/2023
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A0908	Incorrect, Inadequate or Imprecise Result or Readings	07/02/2023	07/02/2023
							F08	Hospitalization or Prolonged Hospitalization	A0908	Incorrect, Inadequate or Imprecise Result or Readings	07/02/2023	07/02/2023
							F1905	Device Revision or Replacement	A0908	Incorrect, Inadequate or Imprecise Result or Readings	07/02/2023	07/02/2023

					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A0908	Incorrect, Inadequate or Imprecise Result or Readings	07/02/2023	07/02/2023
							F08	Hospitalization or Prolonged Hospitalization	A0908	Incorrect, Inadequate or Imprecise Result or Readings	07/02/2023	07/02/2023
							F1905	Device Revision or Replacement	A0908	Incorrect, Inadequate or Imprecise Result or Readings	07/02/2023	07/02/2023
12276	2023	All other reportable incidents	47265	Dual-chamber implantable pacemaker, rate-responsive	CT141	Electrodes/Leads and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	28/11/2022	28/11/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	28/11/2022	28/11/2022
					CT145	Cardiovascular devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	28/11/2022	28/11/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	28/11/2022	28/11/2022
					CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	28/11/2022	28/11/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	28/11/2022	28/11/2022
					CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	28/11/2022	28/11/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	28/11/2022	28/11/2022
					CT233	Surgical	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	28/11/2022	28/11/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	28/11/2022	28/11/2022
					CT2406	Implantable	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	28/11/2022	28/11/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	28/11/2022	28/11/2022
					CT2407	Active implantable	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	28/11/2022	28/11/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	28/11/2022	28/11/2022
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	28/11/2022	28/11/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	28/11/2022	28/11/2022
					CT2727	Active implant devices/ancillaries	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	28/11/2022	28/11/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	28/11/2022	28/11/2022
					CT2750	Implantable cardiac electrical stimulators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	28/11/2022	28/11/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	28/11/2022	28/11/2022
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	28/11/2022	28/11/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	28/11/2022	28/11/2022
					CT2985	Active	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	28/11/2022	28/11/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	28/11/2022	28/11/2022
					CT3001	Implantable metal	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	28/11/2022	28/11/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	28/11/2022	28/11/2022
					CT3002	Implantable synthetic polymer	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	28/11/2022	28/11/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	28/11/2022	28/11/2022
					CT3006	Implantable inorganic material	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	28/11/2022	28/11/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	28/11/2022	28/11/2022
					CT314	Electrically powered	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	28/11/2022	28/11/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	28/11/2022	28/11/2022
					CT321	Long-term surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	28/11/2022	28/11/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	28/11/2022	28/11/2022
					CT449	Generators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	28/11/2022	28/11/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	28/11/2022	28/11/2022
					CT981	Single use	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	28/11/2022	28/11/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	28/11/2022	28/11/2022
					CT983	Surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	28/11/2022	28/11/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	28/11/2022	28/11/2022
12277	2023		15993	Pacemaker programmer	CT1684	Cardiac pacemakers and associated devices	F08	Hospitalization or Prolonged Hospitalization	A0201	Product Quality Problem		
12278	2023	All other reportable incidents	35223	Endocardial/interventricular septal pacing lead	CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070901	Decreased Sensitivity		
							F26	No Health Consequences or Impact	A070901	Decreased Sensitivity		
12279	2023	All other reportable incidents	37265	Dual-chamber implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E060109	Tachycardia	A0709	Device Sensing Problem		
									A071301	Failure to Deliver Shock/Stimulation		
							F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A0709	Device Sensing Problem		
									A071301	Failure to Deliver Shock/Stimulation		
							F1001	Absence of Treatment	A0709	Device Sensing Problem		
									A071301	Failure to Deliver Shock/Stimulation		
							F1203	Life Threatening Illness or Injury	A0709	Device Sensing Problem		
									A071301	Failure to Deliver Shock/Stimulation		
12280	2023		17882	Physiologic-monitoring defibrillation system	CT1684	Cardiac pacemakers and associated devices						
12281	2023	All other reportable incidents	35852	Single-chamber implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A072201	High Impedance		
							F26	No Health Consequences or Impact	A072201	High Impedance		
12282	2023	All other reportable incidents	36241	Pacemaker/defibrillator lead	CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A072201	High Impedance		
							F26	No Health Consequences or Impact	A072201	High Impedance		
12283	2023	All other reportable incidents	35852	Single-chamber implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E2104	Electric Shock	A070909	Oversensing		
									A071302	Inappropriate/Inadequate Shock/Stimulation		
							F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A070909	Oversensing		
									A071302	Inappropriate/Inadequate Shock/Stimulation		
12284	2023		47265	Dual-chamber implantable pacemaker, rate-responsive	CT141	Electrodes/Leads and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	06/02/2023	06/02/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	06/02/2023	06/02/2023
					CT145	Cardiovascular devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	06/02/2023	06/02/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	06/02/2023	06/02/2023
					CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	06/02/2023	06/02/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	06/02/2023	06/02/2023
					CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	06/02/2023	06/02/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	06/02/2023	06/02/2023
					CT233	Surgical	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	06/02/2023	06/02/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	06/02/2023	06/02/2023
					CT2406	Implantable	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	06/02/2023	06/02/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	06/02/2023	06/02/2023
					CT2407	Active implantable	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	06/02/2023	06/02/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	06/02/2023	06/02/2023
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	06/02/2023	06/02/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	06/02/2023	06/02/2023
					CT2727	Active implant devices/ancillaries	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	06/02/2023	06/02/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	06/02/2023	06/02/2023
					CT2750	Implantable cardiac electrical stimulators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	06/02/2023	06/02/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	06/02/2023	06/02/2023
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	06/02/2023	06/02/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	06/02/2023	06/02/2023

					CT2985	Active	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	06/02/2023	06/02/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	06/02/2023	06/02/2023
					CT3001	Implantable metal	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	06/02/2023	06/02/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	06/02/2023	06/02/2023
					CT3002	Implantable synthetic polymer	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	06/02/2023	06/02/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	06/02/2023	06/02/2023
					CT3006	Implantable inorganic material	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	06/02/2023	06/02/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	06/02/2023	06/02/2023
					CT314	Electrically powered	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	06/02/2023	06/02/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	06/02/2023	06/02/2023
					CT321	Long-term surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	06/02/2023	06/02/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	06/02/2023	06/02/2023
					CT449	Generators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	06/02/2023	06/02/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	06/02/2023	06/02/2023
					CT981	Single use	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	06/02/2023	06/02/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	06/02/2023	06/02/2023
					CT983	Surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	06/02/2023	06/02/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	06/02/2023	06/02/2023
12285	2023		47265	Dual-chamber implantable pacemaker, rate-responsive	CT141	Electrodes/Leads and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	02/02/2023	02/02/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	02/02/2023	02/02/2023
					CT145	Cardiovascular devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	02/02/2023	02/02/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	02/02/2023	02/02/2023
					CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	02/02/2023	02/02/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	02/02/2023	02/02/2023
					CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	02/02/2023	02/02/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	02/02/2023	02/02/2023
					CT233	Surgical	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	02/02/2023	02/02/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	02/02/2023	02/02/2023
					CT2406	Implantable	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	02/02/2023	02/02/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	02/02/2023	02/02/2023
					CT2407	Active implantable	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	02/02/2023	02/02/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	02/02/2023	02/02/2023
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	02/02/2023	02/02/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	02/02/2023	02/02/2023
					CT2727	Active implant devices/ancillaries	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	02/02/2023	02/02/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	02/02/2023	02/02/2023
					CT2750	Implantable cardiac electrical stimulators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	02/02/2023	02/02/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	02/02/2023	02/02/2023
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	02/02/2023	02/02/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	02/02/2023	02/02/2023
					CT2985	Active	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	02/02/2023	02/02/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	02/02/2023	02/02/2023
					CT3001	Implantable metal	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	02/02/2023	02/02/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	02/02/2023	02/02/2023
					CT3002	Implantable synthetic polymer	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	02/02/2023	02/02/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	02/02/2023	02/02/2023
					CT3006	Implantable inorganic material	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	02/02/2023	02/02/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	02/02/2023	02/02/2023
					CT314	Electrically powered	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	02/02/2023	02/02/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	02/02/2023	02/02/2023
					CT321	Long-term surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	02/02/2023	02/02/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	02/02/2023	02/02/2023
					CT449	Generators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	02/02/2023	02/02/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	02/02/2023	02/02/2023
					CT981	Single use	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	02/02/2023	02/02/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	02/02/2023	02/02/2023
					CT983	Surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	02/02/2023	02/02/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	02/02/2023	02/02/2023
12286	2023		47265	Dual-chamber implantable pacemaker, rate-responsive	CT141	Electrodes/Leads and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	15/01/2023	15/01/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	15/01/2023	15/01/2023
					CT145	Cardiovascular devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	15/01/2023	15/01/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	15/01/2023	15/01/2023
					CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	15/01/2023	15/01/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	15/01/2023	15/01/2023
					CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	15/01/2023	15/01/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	15/01/2023	15/01/2023
					CT233	Surgical	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	15/01/2023	15/01/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	15/01/2023	15/01/2023
					CT2406	Implantable	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	15/01/2023	15/01/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	15/01/2023	15/01/2023
					CT2407	Active implantable	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	15/01/2023	15/01/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	15/01/2023	15/01/2023
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	15/01/2023	15/01/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	15/01/2023	15/01/2023
					CT2727	Active implant devices/ancillaries	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	15/01/2023	15/01/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	15/01/2023	15/01/2023
					CT2750	Implantable cardiac electrical stimulators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	15/01/2023	15/01/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	15/01/2023	15/01/2023
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	15/01/2023	15/01/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	15/01/2023	15/01/2023
					CT2985	Active	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	15/01/2023	15/01/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	15/01/2023	15/01/2023
					CT3001	Implantable metal	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	15/01/2023	15/01/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	15/01/2023	15/01/2023
					CT3002	Implantable synthetic polymer	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	15/01/2023	15/01/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	15/01/2023	15/01/2023
					CT3006	Implantable inorganic material	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	15/01/2023	15/01/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	15/01/2023	15/01/2023

					CT314	Electrically powered	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	15/01/2023	15/01/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	15/01/2023	15/01/2023
					CT321	Long-term surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	15/01/2023	15/01/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	15/01/2023	15/01/2023
					CT449	Generators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	15/01/2023	15/01/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	15/01/2023	15/01/2023
					CT981	Single use	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	15/01/2023	15/01/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	15/01/2023	15/01/2023
					CT983	Surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	15/01/2023	15/01/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	15/01/2023	15/01/2023
12287	2023		47265	Dual-chamber implantable pacemaker, rate-responsive	CT141	Electrodes/Leads and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	15/01/2023	15/01/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	15/01/2023	15/01/2023
					CT145	Cardiovascular devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	15/01/2023	15/01/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	15/01/2023	15/01/2023
					CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	15/01/2023	15/01/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	15/01/2023	15/01/2023
					CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	15/01/2023	15/01/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	15/01/2023	15/01/2023
					CT233	Surgical	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	15/01/2023	15/01/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	15/01/2023	15/01/2023
					CT2406	Implantable	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	15/01/2023	15/01/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	15/01/2023	15/01/2023
					CT2407	Active implantable	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	15/01/2023	15/01/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	15/01/2023	15/01/2023
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	15/01/2023	15/01/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	15/01/2023	15/01/2023
					CT2727	Active implant devices/ancillaries	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	15/01/2023	15/01/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	15/01/2023	15/01/2023
					CT2750	Implantable cardiac electrical stimulators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	15/01/2023	15/01/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	15/01/2023	15/01/2023
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	15/01/2023	15/01/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	15/01/2023	15/01/2023
					CT2985	Active	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	15/01/2023	15/01/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	15/01/2023	15/01/2023
					CT3001	Implantable metal	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	15/01/2023	15/01/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	15/01/2023	15/01/2023
					CT3002	Implantable synthetic polymer	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	15/01/2023	15/01/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	15/01/2023	15/01/2023
					CT3006	Implantable inorganic material	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	15/01/2023	15/01/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	15/01/2023	15/01/2023
					CT314	Electrically powered	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	15/01/2023	15/01/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	15/01/2023	15/01/2023
					CT321	Long-term surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	15/01/2023	15/01/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	15/01/2023	15/01/2023
					CT449	Generators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	15/01/2023	15/01/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	15/01/2023	15/01/2023
					CT981	Single use	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	15/01/2023	15/01/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	15/01/2023	15/01/2023
					CT983	Surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	15/01/2023	15/01/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	15/01/2023	15/01/2023
12288	2023		47265	Dual-chamber implantable pacemaker, rate-responsive	CT141	Electrodes/Leads and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	11/02/2023	11/02/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	11/02/2023	11/02/2023
					CT145	Cardiovascular devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	11/02/2023	11/02/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	11/02/2023	11/02/2023
					CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	11/02/2023	11/02/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	11/02/2023	11/02/2023
					CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	11/02/2023	11/02/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	11/02/2023	11/02/2023
					CT233	Surgical	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	11/02/2023	11/02/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	11/02/2023	11/02/2023
					CT2406	Implantable	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	11/02/2023	11/02/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	11/02/2023	11/02/2023
					CT2407	Active implantable	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	11/02/2023	11/02/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	11/02/2023	11/02/2023
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	11/02/2023	11/02/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	11/02/2023	11/02/2023
					CT2727	Active implant devices/ancillaries	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	11/02/2023	11/02/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	11/02/2023	11/02/2023
					CT2750	Implantable cardiac electrical stimulators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	11/02/2023	11/02/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	11/02/2023	11/02/2023
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	11/02/2023	11/02/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	11/02/2023	11/02/2023
					CT2985	Active	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	11/02/2023	11/02/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	11/02/2023	11/02/2023
					CT3001	Implantable metal	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	11/02/2023	11/02/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	11/02/2023	11/02/2023
					CT3002	Implantable synthetic polymer	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	11/02/2023	11/02/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	11/02/2023	11/02/2023
					CT3006	Implantable inorganic material	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	11/02/2023	11/02/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	11/02/2023	11/02/2023
					CT314	Electrically powered	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	11/02/2023	11/02/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	11/02/2023	11/02/2023
					CT321	Long-term surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	11/02/2023	11/02/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	11/02/2023	11/02/2023
					CT449	Generators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	11/02/2023	11/02/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	11/02/2023	11/02/2023
					CT981	Single use	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	11/02/2023	11/02/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	11/02/2023	11/02/2023

					CT983	Surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	11/02/2023	11/02/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	11/02/2023	11/02/2023
12289	2023		47265	Dual-chamber implantable pacemaker, rate-responsive	CT141	Electrodes/Leads and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	05/02/2023	05/02/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	05/02/2023	05/02/2023
					CT145	Cardiovascular devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	05/02/2023	05/02/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	05/02/2023	05/02/2023
					CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	05/02/2023	05/02/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	05/02/2023	05/02/2023
					CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	05/02/2023	05/02/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	05/02/2023	05/02/2023
					CT233	Surgical	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	05/02/2023	05/02/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	05/02/2023	05/02/2023
					CT2406	Implantable	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	05/02/2023	05/02/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	05/02/2023	05/02/2023
					CT2407	Active implantable	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	05/02/2023	05/02/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	05/02/2023	05/02/2023
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	05/02/2023	05/02/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	05/02/2023	05/02/2023
					CT2727	Active implant devices/ancillaries	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	05/02/2023	05/02/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	05/02/2023	05/02/2023
					CT2750	Implantable cardiac electrical stimulators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	05/02/2023	05/02/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	05/02/2023	05/02/2023
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	05/02/2023	05/02/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	05/02/2023	05/02/2023
					CT2985	Active	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	05/02/2023	05/02/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	05/02/2023	05/02/2023
					CT3001	Implantable metal	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	05/02/2023	05/02/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	05/02/2023	05/02/2023
					CT3002	Implantable synthetic polymer	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	05/02/2023	05/02/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	05/02/2023	05/02/2023
					CT3006	Implantable inorganic material	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	05/02/2023	05/02/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	05/02/2023	05/02/2023
					CT314	Electrically powered	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	05/02/2023	05/02/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	05/02/2023	05/02/2023
					CT321	Long-term surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	05/02/2023	05/02/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	05/02/2023	05/02/2023
					CT449	Generators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	05/02/2023	05/02/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	05/02/2023	05/02/2023
					CT981	Single use	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	05/02/2023	05/02/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	05/02/2023	05/02/2023
					CT983	Surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	05/02/2023	05/02/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	05/02/2023	05/02/2023
12290	2023		47265	Dual-chamber implantable pacemaker, rate-responsive	CT141	Electrodes/Leads and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	15/01/2023	15/01/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	15/01/2023	15/01/2023
					CT145	Cardiovascular devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	15/01/2023	15/01/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	15/01/2023	15/01/2023
					CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	15/01/2023	15/01/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	15/01/2023	15/01/2023
					CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	15/01/2023	15/01/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	15/01/2023	15/01/2023
					CT233	Surgical	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	15/01/2023	15/01/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	15/01/2023	15/01/2023
					CT2406	Implantable	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	15/01/2023	15/01/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	15/01/2023	15/01/2023
					CT2407	Active implantable	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	15/01/2023	15/01/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	15/01/2023	15/01/2023
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	15/01/2023	15/01/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	15/01/2023	15/01/2023
					CT2727	Active implant devices/ancillaries	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	15/01/2023	15/01/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	15/01/2023	15/01/2023
					CT2750	Implantable cardiac electrical stimulators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	15/01/2023	15/01/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	15/01/2023	15/01/2023
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	15/01/2023	15/01/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	15/01/2023	15/01/2023
					CT2985	Active	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	15/01/2023	15/01/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	15/01/2023	15/01/2023
					CT3001	Implantable metal	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	15/01/2023	15/01/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	15/01/2023	15/01/2023
					CT3002	Implantable synthetic polymer	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	15/01/2023	15/01/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	15/01/2023	15/01/2023
					CT3006	Implantable inorganic material	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	15/01/2023	15/01/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	15/01/2023	15/01/2023
					CT314	Electrically powered	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	15/01/2023	15/01/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	15/01/2023	15/01/2023
					CT321	Long-term surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	15/01/2023	15/01/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	15/01/2023	15/01/2023
					CT449	Generators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	15/01/2023	15/01/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	15/01/2023	15/01/2023
					CT981	Single use	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	15/01/2023	15/01/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	15/01/2023	15/01/2023
					CT983	Surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	15/01/2023	15/01/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	15/01/2023	15/01/2023
12291	2023		47265	Dual-chamber implantable pacemaker, rate-responsive	CT141	Electrodes/Leads and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	31/01/2023	31/01/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	31/01/2023	31/01/2023
					CT145	Cardiovascular devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	31/01/2023	31/01/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	31/01/2023	31/01/2023
					CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	31/01/2023	31/01/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	31/01/2023	31/01/2023

					CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	31/01/2023	31/01/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	31/01/2023	31/01/2023
					CT233	Surgical	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	31/01/2023	31/01/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	31/01/2023	31/01/2023
					CT2406	Implantable	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	31/01/2023	31/01/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	31/01/2023	31/01/2023
					CT2407	Active implantable	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	31/01/2023	31/01/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	31/01/2023	31/01/2023
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	31/01/2023	31/01/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	31/01/2023	31/01/2023
					CT2727	Active implant devices/ancillaries	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	31/01/2023	31/01/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	31/01/2023	31/01/2023
					CT2750	Implantable cardiac electrical stimulators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	31/01/2023	31/01/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	31/01/2023	31/01/2023
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	31/01/2023	31/01/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	31/01/2023	31/01/2023
					CT2985	Active	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	31/01/2023	31/01/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	31/01/2023	31/01/2023
					CT3001	Implantable metal	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	31/01/2023	31/01/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	31/01/2023	31/01/2023
					CT3002	Implantable synthetic polymer	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	31/01/2023	31/01/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	31/01/2023	31/01/2023
					CT3006	Implantable inorganic material	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	31/01/2023	31/01/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	31/01/2023	31/01/2023
					CT314	Electrically powered	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	31/01/2023	31/01/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	31/01/2023	31/01/2023
					CT321	Long-term surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	31/01/2023	31/01/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	31/01/2023	31/01/2023
					CT449	Generators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	31/01/2023	31/01/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	31/01/2023	31/01/2023
					CT981	Single use	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	31/01/2023	31/01/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	31/01/2023	31/01/2023
					CT983	Surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	31/01/2023	31/01/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	31/01/2023	31/01/2023
12292	2023		47265	Dual-chamber implantable pacemaker, rate-responsive	CT141	Electrodes/Leads and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	17/01/2023	17/01/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	17/01/2023	17/01/2023
					CT145	Cardiovascular devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	17/01/2023	17/01/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	17/01/2023	17/01/2023
					CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	17/01/2023	17/01/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	17/01/2023	17/01/2023
					CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	17/01/2023	17/01/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	17/01/2023	17/01/2023
					CT233	Surgical	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	17/01/2023	17/01/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	17/01/2023	17/01/2023
					CT2406	Implantable	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	17/01/2023	17/01/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	17/01/2023	17/01/2023
					CT2407	Active implantable	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	17/01/2023	17/01/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	17/01/2023	17/01/2023
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	17/01/2023	17/01/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	17/01/2023	17/01/2023
					CT2727	Active implant devices/ancillaries	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	17/01/2023	17/01/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	17/01/2023	17/01/2023
					CT2750	Implantable cardiac electrical stimulators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	17/01/2023	17/01/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	17/01/2023	17/01/2023
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	17/01/2023	17/01/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	17/01/2023	17/01/2023
					CT2985	Active	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	17/01/2023	17/01/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	17/01/2023	17/01/2023
					CT3001	Implantable metal	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	17/01/2023	17/01/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	17/01/2023	17/01/2023
					CT3002	Implantable synthetic polymer	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	17/01/2023	17/01/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	17/01/2023	17/01/2023
					CT3006	Implantable inorganic material	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	17/01/2023	17/01/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	17/01/2023	17/01/2023
					CT314	Electrically powered	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	17/01/2023	17/01/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	17/01/2023	17/01/2023
					CT321	Long-term surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	17/01/2023	17/01/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	17/01/2023	17/01/2023
					CT449	Generators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	17/01/2023	17/01/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	17/01/2023	17/01/2023
					CT981	Single use	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	17/01/2023	17/01/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	17/01/2023	17/01/2023
					CT983	Surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	17/01/2023	17/01/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	17/01/2023	17/01/2023
12293	2023		47265	Dual-chamber implantable pacemaker, rate-responsive	CT141	Electrodes/Leads and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	15/01/2023	15/01/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	15/01/2023	15/01/2023
					CT145	Cardiovascular devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	15/01/2023	15/01/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	15/01/2023	15/01/2023
					CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	15/01/2023	15/01/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	15/01/2023	15/01/2023
					CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	15/01/2023	15/01/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	15/01/2023	15/01/2023
					CT233	Surgical	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	15/01/2023	15/01/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	15/01/2023	15/01/2023
					CT2406	Implantable	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	15/01/2023	15/01/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	15/01/2023	15/01/2023
					CT2407	Active implantable	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	15/01/2023	15/01/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	15/01/2023	15/01/2023

					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	15/01/2023	15/01/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	15/01/2023	15/01/2023
					CT2727	Active implant devices/ancillaries	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	15/01/2023	15/01/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	15/01/2023	15/01/2023
					CT2750	Implantable cardiac electrical stimulators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	15/01/2023	15/01/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	15/01/2023	15/01/2023
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	15/01/2023	15/01/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	15/01/2023	15/01/2023
					CT2985	Active	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	15/01/2023	15/01/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	15/01/2023	15/01/2023
					CT3001	Implantable metal	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	15/01/2023	15/01/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	15/01/2023	15/01/2023
					CT3002	Implantable synthetic polymer	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	15/01/2023	15/01/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	15/01/2023	15/01/2023
					CT3006	Implantable inorganic material	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	15/01/2023	15/01/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	15/01/2023	15/01/2023
					CT314	Electrically powered	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	15/01/2023	15/01/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	15/01/2023	15/01/2023
					CT321	Long-term surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	15/01/2023	15/01/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	15/01/2023	15/01/2023
					CT449	Generators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	15/01/2023	15/01/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	15/01/2023	15/01/2023
					CT981	Single use	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	15/01/2023	15/01/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	15/01/2023	15/01/2023
					CT983	Surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	15/01/2023	15/01/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	15/01/2023	15/01/2023
12294	2023	All other reportable incidents	47265	Dual-chamber implantable pacemaker, rate-responsive	CT141	Electrodes/Leads and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	24/01/2023	24/01/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	24/01/2023	24/01/2023
					CT145	Cardiovascular devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	24/01/2023	24/01/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	24/01/2023	24/01/2023
					CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	24/01/2023	24/01/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	24/01/2023	24/01/2023
					CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	24/01/2023	24/01/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	24/01/2023	24/01/2023
					CT233	Surgical	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	24/01/2023	24/01/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	24/01/2023	24/01/2023
					CT2406	Implantable	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	24/01/2023	24/01/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	24/01/2023	24/01/2023
					CT2407	Active implantable	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	24/01/2023	24/01/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	24/01/2023	24/01/2023
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	24/01/2023	24/01/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	24/01/2023	24/01/2023
					CT2727	Active implant devices/ancillaries	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	24/01/2023	24/01/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	24/01/2023	24/01/2023
					CT2750	Implantable cardiac electrical stimulators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	24/01/2023	24/01/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	24/01/2023	24/01/2023
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	24/01/2023	24/01/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	24/01/2023	24/01/2023
					CT2985	Active	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	24/01/2023	24/01/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	24/01/2023	24/01/2023
					CT3001	Implantable metal	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	24/01/2023	24/01/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	24/01/2023	24/01/2023
					CT3002	Implantable synthetic polymer	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	24/01/2023	24/01/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	24/01/2023	24/01/2023
					CT3006	Implantable inorganic material	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	24/01/2023	24/01/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	24/01/2023	24/01/2023
					CT314	Electrically powered	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	24/01/2023	24/01/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	24/01/2023	24/01/2023
					CT321	Long-term surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	24/01/2023	24/01/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	24/01/2023	24/01/2023
					CT449	Generators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	24/01/2023	24/01/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	24/01/2023	24/01/2023
					CT981	Single use	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	24/01/2023	24/01/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	24/01/2023	24/01/2023
					CT983	Surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	24/01/2023	24/01/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	24/01/2023	24/01/2023
12295	2023	All other reportable incidents	47265	Dual-chamber implantable pacemaker, rate-responsive	CT141	Electrodes/Leads and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	24/01/2023	24/01/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	24/01/2023	24/01/2023
					CT145	Cardiovascular devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	24/01/2023	24/01/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	24/01/2023	24/01/2023
					CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	24/01/2023	24/01/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	24/01/2023	24/01/2023
					CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	24/01/2023	24/01/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	24/01/2023	24/01/2023
					CT233	Surgical	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	24/01/2023	24/01/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	24/01/2023	24/01/2023
					CT2406	Implantable	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	24/01/2023	24/01/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	24/01/2023	24/01/2023
					CT2407	Active implantable	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	24/01/2023	24/01/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	24/01/2023	24/01/2023
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	24/01/2023	24/01/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	24/01/2023	24/01/2023
					CT2727	Active implant devices/ancillaries	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	24/01/2023	24/01/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	24/01/2023	24/01/2023
					CT2750	Implantable cardiac electrical stimulators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	24/01/2023	24/01/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	24/01/2023	24/01/2023
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	24/01/2023	24/01/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	24/01/2023	24/01/2023

					CT2985	Active	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	24/01/2023	24/01/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	24/01/2023	24/01/2023
					CT3001	Implantable metal	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	24/01/2023	24/01/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	24/01/2023	24/01/2023
					CT3002	Implantable synthetic polymer	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	24/01/2023	24/01/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	24/01/2023	24/01/2023
					CT3006	Implantable inorganic material	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	24/01/2023	24/01/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	24/01/2023	24/01/2023
					CT314	Electrically powered	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	24/01/2023	24/01/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	24/01/2023	24/01/2023
					CT321	Long-term surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	24/01/2023	24/01/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	24/01/2023	24/01/2023
					CT449	Generators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	24/01/2023	24/01/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	24/01/2023	24/01/2023
					CT981	Single use	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	24/01/2023	24/01/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	24/01/2023	24/01/2023
					CT983	Surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	24/01/2023	24/01/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	24/01/2023	24/01/2023
12296	2023	All other reportable incidents	35223	Endocardial/interventricular septal pacing lead	CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070909	Oversensing		
							F26	No Health Consequences or Impact	A070909	Oversensing		
12297	2023	All other reportable incidents	47265	Dual-chamber implantable pacemaker, rate-responsive	CT141	Electrodes/Leads and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	09/11/2022	09/11/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	09/11/2022	09/11/2022
					CT145	Cardiovascular devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	09/11/2022	09/11/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	09/11/2022	09/11/2022
					CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	09/11/2022	09/11/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	09/11/2022	09/11/2022
					CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	09/11/2022	09/11/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	09/11/2022	09/11/2022
					CT233	Surgical	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	09/11/2022	09/11/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	09/11/2022	09/11/2022
					CT2406	Implantable	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	09/11/2022	09/11/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	09/11/2022	09/11/2022
					CT2407	Active implantable	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	09/11/2022	09/11/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	09/11/2022	09/11/2022
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	09/11/2022	09/11/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	09/11/2022	09/11/2022
					CT2727	Active implant devices/ancillaries	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	09/11/2022	09/11/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	09/11/2022	09/11/2022
					CT2750	Implantable cardiac electrical stimulators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	09/11/2022	09/11/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	09/11/2022	09/11/2022
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	09/11/2022	09/11/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	09/11/2022	09/11/2022
					CT2985	Active	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	09/11/2022	09/11/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	09/11/2022	09/11/2022
					CT3001	Implantable metal	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	09/11/2022	09/11/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	09/11/2022	09/11/2022
					CT3002	Implantable synthetic polymer	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	09/11/2022	09/11/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	09/11/2022	09/11/2022
					CT3006	Implantable inorganic material	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	09/11/2022	09/11/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	09/11/2022	09/11/2022
					CT314	Electrically powered	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	09/11/2022	09/11/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	09/11/2022	09/11/2022
					CT321	Long-term surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	09/11/2022	09/11/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	09/11/2022	09/11/2022
					CT449	Generators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	09/11/2022	09/11/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	09/11/2022	09/11/2022
					CT981	Single use	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	09/11/2022	09/11/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	09/11/2022	09/11/2022
					CT983	Surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	09/11/2022	09/11/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	09/11/2022	09/11/2022
12298	2023		47265	Dual-chamber implantable pacemaker, rate-responsive	CT141	Electrodes/Leads and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	22/01/2023	22/01/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	22/01/2023	22/01/2023
					CT145	Cardiovascular devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	22/01/2023	22/01/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	22/01/2023	22/01/2023
					CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	22/01/2023	22/01/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	22/01/2023	22/01/2023
					CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	22/01/2023	22/01/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	22/01/2023	22/01/2023
					CT233	Surgical	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	22/01/2023	22/01/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	22/01/2023	22/01/2023
					CT2406	Implantable	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	22/01/2023	22/01/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	22/01/2023	22/01/2023
					CT2407	Active implantable	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	22/01/2023	22/01/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	22/01/2023	22/01/2023
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	22/01/2023	22/01/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	22/01/2023	22/01/2023
					CT2727	Active implant devices/ancillaries	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	22/01/2023	22/01/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	22/01/2023	22/01/2023
					CT2750	Implantable cardiac electrical stimulators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	22/01/2023	22/01/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	22/01/2023	22/01/2023
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	22/01/2023	22/01/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	22/01/2023	22/01/2023
					CT2985	Active	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	22/01/2023	22/01/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	22/01/2023	22/01/2023
					CT3001	Implantable metal	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	22/01/2023	22/01/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	22/01/2023	22/01/2023
					CT3002	Implantable synthetic polymer	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	22/01/2023	22/01/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	22/01/2023	22/01/2023

					CT3006	Implantable inorganic material	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	22/01/2023	22/01/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	22/01/2023	22/01/2023
					CT314	Electrically powered	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	22/01/2023	22/01/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	22/01/2023	22/01/2023
					CT321	Long-term surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	22/01/2023	22/01/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	22/01/2023	22/01/2023
					CT449	Generators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	22/01/2023	22/01/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	22/01/2023	22/01/2023
					CT981	Single use	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	22/01/2023	22/01/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	22/01/2023	22/01/2023
					CT983	Surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	22/01/2023	22/01/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	22/01/2023	22/01/2023
12299	2023		47265	Dual-chamber implantable pacemaker, rate-responsive	CT141	Electrodes/Leads and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	15/01/2023	15/01/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	15/01/2023	15/01/2023
					CT145	Cardiovascular devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	15/01/2023	15/01/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	15/01/2023	15/01/2023
					CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	15/01/2023	15/01/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	15/01/2023	15/01/2023
					CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	15/01/2023	15/01/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	15/01/2023	15/01/2023
					CT233	Surgical	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	15/01/2023	15/01/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	15/01/2023	15/01/2023
					CT2406	Implantable	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	15/01/2023	15/01/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	15/01/2023	15/01/2023
					CT2407	Active implantable	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	15/01/2023	15/01/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	15/01/2023	15/01/2023
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	15/01/2023	15/01/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	15/01/2023	15/01/2023
					CT2727	Active implant devices/ancillaries	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	15/01/2023	15/01/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	15/01/2023	15/01/2023
					CT2750	Implantable cardiac electrical stimulators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	15/01/2023	15/01/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	15/01/2023	15/01/2023
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	15/01/2023	15/01/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	15/01/2023	15/01/2023
					CT2985	Active	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	15/01/2023	15/01/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	15/01/2023	15/01/2023
					CT3001	Implantable metal	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	15/01/2023	15/01/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	15/01/2023	15/01/2023
					CT3002	Implantable synthetic polymer	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	15/01/2023	15/01/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	15/01/2023	15/01/2023
					CT3006	Implantable inorganic material	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	15/01/2023	15/01/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	15/01/2023	15/01/2023
					CT314	Electrically powered	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	15/01/2023	15/01/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	15/01/2023	15/01/2023
					CT321	Long-term surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	15/01/2023	15/01/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	15/01/2023	15/01/2023
					CT449	Generators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	15/01/2023	15/01/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	15/01/2023	15/01/2023
					CT981	Single use	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	15/01/2023	15/01/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	15/01/2023	15/01/2023
					CT983	Surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	15/01/2023	15/01/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	15/01/2023	15/01/2023
12300	2023		47265	Dual-chamber implantable pacemaker, rate-responsive	CT141	Electrodes/Leads and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	15/01/2023	15/01/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	15/01/2023	15/01/2023
					CT145	Cardiovascular devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	15/01/2023	15/01/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	15/01/2023	15/01/2023
					CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	15/01/2023	15/01/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	15/01/2023	15/01/2023
					CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	15/01/2023	15/01/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	15/01/2023	15/01/2023
					CT233	Surgical	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	15/01/2023	15/01/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	15/01/2023	15/01/2023
					CT2406	Implantable	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	15/01/2023	15/01/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	15/01/2023	15/01/2023
					CT2407	Active implantable	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	15/01/2023	15/01/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	15/01/2023	15/01/2023
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	15/01/2023	15/01/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	15/01/2023	15/01/2023
					CT2727	Active implant devices/ancillaries	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	15/01/2023	15/01/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	15/01/2023	15/01/2023
					CT2750	Implantable cardiac electrical stimulators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	15/01/2023	15/01/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	15/01/2023	15/01/2023
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	15/01/2023	15/01/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	15/01/2023	15/01/2023
					CT2985	Active	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	15/01/2023	15/01/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	15/01/2023	15/01/2023
					CT3001	Implantable metal	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	15/01/2023	15/01/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	15/01/2023	15/01/2023
					CT3002	Implantable synthetic polymer	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	15/01/2023	15/01/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	15/01/2023	15/01/2023
					CT3006	Implantable inorganic material	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	15/01/2023	15/01/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	15/01/2023	15/01/2023
					CT314	Electrically powered	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	15/01/2023	15/01/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	15/01/2023	15/01/2023
					CT321	Long-term surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	15/01/2023	15/01/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	15/01/2023	15/01/2023
					CT449	Generators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	15/01/2023	15/01/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	15/01/2023	15/01/2023

					CT981	Single use	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	15/01/2023	15/01/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	15/01/2023	15/01/2023
					CT983	Surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	15/01/2023	15/01/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	15/01/2023	15/01/2023
12301	2023		47265	Dual-chamber implantable pacemaker, rate-responsive	CT141	Electrodes/Leads and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	15/01/2023	15/01/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	15/01/2023	15/01/2023
					CT145	Cardiovascular devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	15/01/2023	15/01/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	15/01/2023	15/01/2023
					CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	15/01/2023	15/01/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	15/01/2023	15/01/2023
					CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	15/01/2023	15/01/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	15/01/2023	15/01/2023
					CT233	Surgical	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	15/01/2023	15/01/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	15/01/2023	15/01/2023
					CT2406	Implantable	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	15/01/2023	15/01/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	15/01/2023	15/01/2023
					CT2407	Active implantable	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	15/01/2023	15/01/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	15/01/2023	15/01/2023
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	15/01/2023	15/01/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	15/01/2023	15/01/2023
					CT2727	Active implant devices/ancillaries	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	15/01/2023	15/01/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	15/01/2023	15/01/2023
					CT2750	Implantable cardiac electrical stimulators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	15/01/2023	15/01/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	15/01/2023	15/01/2023
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	15/01/2023	15/01/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	15/01/2023	15/01/2023
					CT2985	Active	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	15/01/2023	15/01/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	15/01/2023	15/01/2023
					CT3001	Implantable metal	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	15/01/2023	15/01/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	15/01/2023	15/01/2023
					CT3002	Implantable synthetic polymer	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	15/01/2023	15/01/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	15/01/2023	15/01/2023
					CT3006	Implantable inorganic material	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	15/01/2023	15/01/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	15/01/2023	15/01/2023
					CT314	Electrically powered	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	15/01/2023	15/01/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	15/01/2023	15/01/2023
					CT321	Long-term surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	15/01/2023	15/01/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	15/01/2023	15/01/2023
					CT449	Generators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	15/01/2023	15/01/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	15/01/2023	15/01/2023
					CT981	Single use	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	15/01/2023	15/01/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	15/01/2023	15/01/2023
					CT983	Surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	15/01/2023	15/01/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	15/01/2023	15/01/2023
12302	2023	All other reportable incidents	47270	Cardiac resynchronization therapy implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E2336	Shock	A0712	Pacing Problem		
							F26	No Health Consequences or Impact	A0712	Pacing Problem		
12303	2023	All other reportable incidents	47263	Cardiac resynchronization therapy implantable pacemaker	CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A0908	Incorrect, Inadequate or Imprecise Result or Readings	08/11/2022	08/11/2022
							F08	Hospitalization or Prolonged Hospitalization	A0908	Incorrect, Inadequate or Imprecise Result or Readings	08/11/2022	08/11/2022
							F1905	Device Revision or Replacement	A0908	Incorrect, Inadequate or Imprecise Result or Readings	08/11/2022	08/11/2022
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A0908	Incorrect, Inadequate or Imprecise Result or Readings	08/11/2022	08/11/2022
							F08	Hospitalization or Prolonged Hospitalization	A0908	Incorrect, Inadequate or Imprecise Result or Readings	08/11/2022	08/11/2022
							F1905	Device Revision or Replacement	A0908	Incorrect, Inadequate or Imprecise Result or Readings	08/11/2022	08/11/2022
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A0908	Incorrect, Inadequate or Imprecise Result or Readings	08/11/2022	08/11/2022
							F08	Hospitalization or Prolonged Hospitalization	A0908	Incorrect, Inadequate or Imprecise Result or Readings	08/11/2022	08/11/2022
							F1905	Device Revision or Replacement	A0908	Incorrect, Inadequate or Imprecise Result or Readings	08/11/2022	08/11/2022
12304	2023	All other reportable incidents	47265	Dual-chamber implantable pacemaker, rate-responsive	CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
12305	2023	All other reportable incidents	47265	Dual-chamber implantable pacemaker, rate-responsive	CT1685	Cardiac pacemakers	E0112	Dizziness	A070101	Failure to Capture		
									A0712	Pacing Problem		
							F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A070101	Failure to Capture		
									A0712	Pacing Problem		
					CT257	Implantable cardiac pacemakers	E0112	Dizziness	A070101	Failure to Capture		
									A0712	Pacing Problem		
							F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A070101	Failure to Capture		
									A0712	Pacing Problem		
					CT2866	Implantable cardiac electrical stimulators	E0112	Dizziness	A070101	Failure to Capture		
									A0712	Pacing Problem		
							F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A070101	Failure to Capture		
									A0712	Pacing Problem		
12306	2023	All other reportable incidents	47265	Dual-chamber implantable pacemaker, rate-responsive	CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		

							F26	No Health Consequences or Impact	A0705	Battery Problem			
									A070504	Premature Discharge of Battery			
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem			
									A070504	Premature Discharge of Battery			
							F26	No Health Consequences or Impact	A0705	Battery Problem			
									A070504	Premature Discharge of Battery			
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem			
									A070504	Premature Discharge of Battery			
							F26	No Health Consequences or Impact	A0705	Battery Problem			
									A070504	Premature Discharge of Battery			
12307	2023	All other reportable incidents	47270	Cardiac resynchronization therapy implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	23/01/2023	23/01/2023	
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	23/01/2023	23/01/2023	
							F12	Serious Injury/ Illness/ Impairment	A070504	Premature Discharge of Battery	23/01/2023	23/01/2023	
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	23/01/2023	23/01/2023	
12308	2023		47265	Dual-chamber implantable pacemaker, rate-responsive	CT141	Electrodes/Leads and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	10/12/2022	10/12/2022	
							F1901	Additional Surgery	A24	No Apparent Device Problem	10/12/2022	10/12/2022	
					CT145	Cardiovascular devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	10/12/2022	10/12/2022	
							F1901	Additional Surgery	A24	No Apparent Device Problem	10/12/2022	10/12/2022	
					CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	10/12/2022	10/12/2022	
							F1901	Additional Surgery	A24	No Apparent Device Problem	10/12/2022	10/12/2022	
					CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	10/12/2022	10/12/2022	
							F1901	Additional Surgery	A24	No Apparent Device Problem	10/12/2022	10/12/2022	
					CT233	Surgical	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	10/12/2022	10/12/2022	
							F1901	Additional Surgery	A24	No Apparent Device Problem	10/12/2022	10/12/2022	
					CT2406	Implantable	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	10/12/2022	10/12/2022	
							F1901	Additional Surgery	A24	No Apparent Device Problem	10/12/2022	10/12/2022	
					CT2407	Active implantable	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	10/12/2022	10/12/2022	
							F1901	Additional Surgery	A24	No Apparent Device Problem	10/12/2022	10/12/2022	
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	10/12/2022	10/12/2022	
							F1901	Additional Surgery	A24	No Apparent Device Problem	10/12/2022	10/12/2022	
					CT2727	Active implant devices/ancillaries	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	10/12/2022	10/12/2022	
							F1901	Additional Surgery	A24	No Apparent Device Problem	10/12/2022	10/12/2022	
					CT2750	Implantable cardiac electrical stimulators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	10/12/2022	10/12/2022	
							F1901	Additional Surgery	A24	No Apparent Device Problem	10/12/2022	10/12/2022	
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	10/12/2022	10/12/2022	
							F1901	Additional Surgery	A24	No Apparent Device Problem	10/12/2022	10/12/2022	
					CT2985	Active	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	10/12/2022	10/12/2022	
							F1901	Additional Surgery	A24	No Apparent Device Problem	10/12/2022	10/12/2022	
					CT3001	Implantable metal	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	10/12/2022	10/12/2022	
							F1901	Additional Surgery	A24	No Apparent Device Problem	10/12/2022	10/12/2022	
					CT3002	Implantable synthetic polymer	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	10/12/2022	10/12/2022	
							F1901	Additional Surgery	A24	No Apparent Device Problem	10/12/2022	10/12/2022	
					CT3006	Implantable inorganic material	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	10/12/2022	10/12/2022	
							F1901	Additional Surgery	A24	No Apparent Device Problem	10/12/2022	10/12/2022	
					CT314	Electrically powered	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	10/12/2022	10/12/2022	
							F1901	Additional Surgery	A24	No Apparent Device Problem	10/12/2022	10/12/2022	
					CT321	Long-term surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	10/12/2022	10/12/2022	
							F1901	Additional Surgery	A24	No Apparent Device Problem	10/12/2022	10/12/2022	
					CT449	Generators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	10/12/2022	10/12/2022	
							F1901	Additional Surgery	A24	No Apparent Device Problem	10/12/2022	10/12/2022	
					CT981	Single use	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	10/12/2022	10/12/2022	
							F1901	Additional Surgery	A24	No Apparent Device Problem	10/12/2022	10/12/2022	
					CT983	Surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	10/12/2022	10/12/2022	
							F1901	Additional Surgery	A24	No Apparent Device Problem	10/12/2022	10/12/2022	
12309	2023	All other reportable incidents	47265	Dual-chamber implantable pacemaker, rate-responsive	CT141	Electrodes/Leads and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery			
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery			
					CT145	Cardiovascular devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery			
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery			
					CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery			
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery			
					CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery			
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery			
					CT233	Surgical	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery			
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery			
					CT2406	Implantable	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery			
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery			
					CT2407	Active implantable	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery			
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery			
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery			
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery			
					CT2727	Active implant devices/ancillaries	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery			
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery			
					CT2750	Implantable cardiac electrical stimulators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery			
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery			
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery			
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery			
					CT2985	Active	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery			
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery			
					CT3001	Implantable metal	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery			
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery			
					CT3002	Implantable synthetic polymer	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery			
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery			
					CT3006	Implantable inorganic material	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery			
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery			
					CT314	Electrically powered	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery			
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery			
					CT321	Long-term surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery			
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery			

					CT449	Generators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT981	Single use	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT983	Surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
12310	2023		47265	Dual-chamber implantable pacemaker, rate-responsive	CT141	Electrodes/Leads and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	10/12/2022	10/12/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	10/12/2022	10/12/2022
					CT145	Cardiovascular devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	10/12/2022	10/12/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	10/12/2022	10/12/2022
					CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	10/12/2022	10/12/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	10/12/2022	10/12/2022
					CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	10/12/2022	10/12/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	10/12/2022	10/12/2022
					CT233	Surgical	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	10/12/2022	10/12/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	10/12/2022	10/12/2022
					CT2406	Implantable	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	10/12/2022	10/12/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	10/12/2022	10/12/2022
					CT2407	Active implantable	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	10/12/2022	10/12/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	10/12/2022	10/12/2022
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	10/12/2022	10/12/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	10/12/2022	10/12/2022
					CT2727	Active implant devices/ancillaries	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	10/12/2022	10/12/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	10/12/2022	10/12/2022
					CT2750	Implantable cardiac electrical stimulators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	10/12/2022	10/12/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	10/12/2022	10/12/2022
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	10/12/2022	10/12/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	10/12/2022	10/12/2022
					CT2985	Active	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	10/12/2022	10/12/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	10/12/2022	10/12/2022
					CT3001	Implantable metal	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	10/12/2022	10/12/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	10/12/2022	10/12/2022
					CT3002	Implantable synthetic polymer	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	10/12/2022	10/12/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	10/12/2022	10/12/2022
					CT3006	Implantable inorganic material	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	10/12/2022	10/12/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	10/12/2022	10/12/2022
					CT314	Electrically powered	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	10/12/2022	10/12/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	10/12/2022	10/12/2022
					CT321	Long-term surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	10/12/2022	10/12/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	10/12/2022	10/12/2022
					CT449	Generators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	10/12/2022	10/12/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	10/12/2022	10/12/2022
					CT981	Single use	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	10/12/2022	10/12/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	10/12/2022	10/12/2022
					CT983	Surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	10/12/2022	10/12/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	10/12/2022	10/12/2022
12311	2023		47265	Dual-chamber implantable pacemaker, rate-responsive	CT141	Electrodes/Leads and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	30/01/2023	30/01/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	30/01/2023	30/01/2023
					CT145	Cardiovascular devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	30/01/2023	30/01/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	30/01/2023	30/01/2023
					CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	30/01/2023	30/01/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	30/01/2023	30/01/2023
					CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	30/01/2023	30/01/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	30/01/2023	30/01/2023
					CT233	Surgical	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	30/01/2023	30/01/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	30/01/2023	30/01/2023
					CT2406	Implantable	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	30/01/2023	30/01/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	30/01/2023	30/01/2023
					CT2407	Active implantable	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	30/01/2023	30/01/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	30/01/2023	30/01/2023
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	30/01/2023	30/01/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	30/01/2023	30/01/2023
					CT2727	Active implant devices/ancillaries	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	30/01/2023	30/01/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	30/01/2023	30/01/2023
					CT2750	Implantable cardiac electrical stimulators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	30/01/2023	30/01/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	30/01/2023	30/01/2023
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	30/01/2023	30/01/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	30/01/2023	30/01/2023
					CT2985	Active	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	30/01/2023	30/01/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	30/01/2023	30/01/2023
					CT3001	Implantable metal	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	30/01/2023	30/01/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	30/01/2023	30/01/2023
					CT3002	Implantable synthetic polymer	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	30/01/2023	30/01/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	30/01/2023	30/01/2023
					CT3006	Implantable inorganic material	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	30/01/2023	30/01/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	30/01/2023	30/01/2023
					CT314	Electrically powered	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	30/01/2023	30/01/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	30/01/2023	30/01/2023
					CT321	Long-term surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	30/01/2023	30/01/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	30/01/2023	30/01/2023
					CT449	Generators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	30/01/2023	30/01/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	30/01/2023	30/01/2023
					CT981	Single use	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	30/01/2023	30/01/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	30/01/2023	30/01/2023
					CT983	Surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	30/01/2023	30/01/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	30/01/2023	30/01/2023
12312	2023	All other reportable incidents	47265	Dual-chamber implantable pacemaker, rate-responsive	CT141	Electrodes/Leads and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	03/02/2023	03/02/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	03/02/2023	03/02/2023

					CT145	Cardiovascular devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	03/02/2023	03/02/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	03/02/2023	03/02/2023
					CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	03/02/2023	03/02/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	03/02/2023	03/02/2023
					CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	03/02/2023	03/02/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	03/02/2023	03/02/2023
					CT233	Surgical	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	03/02/2023	03/02/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	03/02/2023	03/02/2023
					CT2406	Implantable	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	03/02/2023	03/02/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	03/02/2023	03/02/2023
					CT2407	Active implantable	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	03/02/2023	03/02/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	03/02/2023	03/02/2023
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	03/02/2023	03/02/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	03/02/2023	03/02/2023
					CT2727	Active implant devices/ancillaries	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	03/02/2023	03/02/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	03/02/2023	03/02/2023
					CT2750	Implantable cardiac electrical stimulators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	03/02/2023	03/02/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	03/02/2023	03/02/2023
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	03/02/2023	03/02/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	03/02/2023	03/02/2023
					CT2985	Active	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	03/02/2023	03/02/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	03/02/2023	03/02/2023
					CT3001	Implantable metal	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	03/02/2023	03/02/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	03/02/2023	03/02/2023
					CT3002	Implantable synthetic polymer	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	03/02/2023	03/02/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	03/02/2023	03/02/2023
					CT3006	Implantable inorganic material	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	03/02/2023	03/02/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	03/02/2023	03/02/2023
					CT314	Electrically powered	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	03/02/2023	03/02/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	03/02/2023	03/02/2023
					CT321	Long-term surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	03/02/2023	03/02/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	03/02/2023	03/02/2023
					CT449	Generators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	03/02/2023	03/02/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	03/02/2023	03/02/2023
					CT981	Single use	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	03/02/2023	03/02/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	03/02/2023	03/02/2023
					CT983	Surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	03/02/2023	03/02/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	03/02/2023	03/02/2023
12313	2023		47265	Dual-chamber implantable pacemaker, rate-responsive	CT141	Electrodes/Leads and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	15/01/2023	15/01/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	15/01/2023	15/01/2023
					CT145	Cardiovascular devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	15/01/2023	15/01/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	15/01/2023	15/01/2023
					CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	15/01/2023	15/01/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	15/01/2023	15/01/2023
					CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	15/01/2023	15/01/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	15/01/2023	15/01/2023
					CT233	Surgical	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	15/01/2023	15/01/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	15/01/2023	15/01/2023
					CT2406	Implantable	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	15/01/2023	15/01/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	15/01/2023	15/01/2023
					CT2407	Active implantable	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	15/01/2023	15/01/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	15/01/2023	15/01/2023
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	15/01/2023	15/01/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	15/01/2023	15/01/2023
					CT2727	Active implant devices/ancillaries	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	15/01/2023	15/01/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	15/01/2023	15/01/2023
					CT2750	Implantable cardiac electrical stimulators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	15/01/2023	15/01/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	15/01/2023	15/01/2023
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	15/01/2023	15/01/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	15/01/2023	15/01/2023
					CT2985	Active	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	15/01/2023	15/01/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	15/01/2023	15/01/2023
					CT3001	Implantable metal	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	15/01/2023	15/01/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	15/01/2023	15/01/2023
					CT3002	Implantable synthetic polymer	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	15/01/2023	15/01/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	15/01/2023	15/01/2023
					CT3006	Implantable inorganic material	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	15/01/2023	15/01/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	15/01/2023	15/01/2023
					CT314	Electrically powered	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	15/01/2023	15/01/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	15/01/2023	15/01/2023
					CT321	Long-term surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	15/01/2023	15/01/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	15/01/2023	15/01/2023
					CT449	Generators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	15/01/2023	15/01/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	15/01/2023	15/01/2023
					CT981	Single use	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	15/01/2023	15/01/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	15/01/2023	15/01/2023
					CT983	Surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	15/01/2023	15/01/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	15/01/2023	15/01/2023
12314	2023		47265	Dual-chamber implantable pacemaker, rate-responsive	CT141	Electrodes/Leads and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	19/01/2023	19/01/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	19/01/2023	19/01/2023
					CT145	Cardiovascular devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	19/01/2023	19/01/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	19/01/2023	19/01/2023
					CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	19/01/2023	19/01/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	19/01/2023	19/01/2023
					CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	19/01/2023	19/01/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	19/01/2023	19/01/2023
					CT233	Surgical	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	19/01/2023	19/01/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	19/01/2023	19/01/2023

					CT2406	Implantable	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	19/01/2023	19/01/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	19/01/2023	19/01/2023
					CT2407	Active implantable	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	19/01/2023	19/01/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	19/01/2023	19/01/2023
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	19/01/2023	19/01/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	19/01/2023	19/01/2023
					CT2727	Active implant devices/ancillaries	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	19/01/2023	19/01/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	19/01/2023	19/01/2023
					CT2750	Implantable cardiac electrical stimulators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	19/01/2023	19/01/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	19/01/2023	19/01/2023
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	19/01/2023	19/01/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	19/01/2023	19/01/2023
					CT2985	Active	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	19/01/2023	19/01/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	19/01/2023	19/01/2023
					CT3001	Implantable metal	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	19/01/2023	19/01/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	19/01/2023	19/01/2023
					CT3002	Implantable synthetic polymer	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	19/01/2023	19/01/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	19/01/2023	19/01/2023
					CT3006	Implantable inorganic material	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	19/01/2023	19/01/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	19/01/2023	19/01/2023
					CT314	Electrically powered	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	19/01/2023	19/01/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	19/01/2023	19/01/2023
					CT321	Long-term surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	19/01/2023	19/01/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	19/01/2023	19/01/2023
					CT449	Generators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	19/01/2023	19/01/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	19/01/2023	19/01/2023
					CT981	Single use	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	19/01/2023	19/01/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	19/01/2023	19/01/2023
					CT983	Surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	19/01/2023	19/01/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	19/01/2023	19/01/2023
12315	2023		47265	Dual-chamber implantable pacemaker, rate-responsive	CT141	Electrodes/Leads and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	31/01/2023	31/01/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	31/01/2023	31/01/2023
					CT145	Cardiovascular devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	31/01/2023	31/01/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	31/01/2023	31/01/2023
					CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	31/01/2023	31/01/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	31/01/2023	31/01/2023
					CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	31/01/2023	31/01/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	31/01/2023	31/01/2023
					CT233	Surgical	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	31/01/2023	31/01/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	31/01/2023	31/01/2023
					CT2406	Implantable	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	31/01/2023	31/01/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	31/01/2023	31/01/2023
					CT2407	Active implantable	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	31/01/2023	31/01/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	31/01/2023	31/01/2023
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	31/01/2023	31/01/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	31/01/2023	31/01/2023
					CT2727	Active implant devices/ancillaries	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	31/01/2023	31/01/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	31/01/2023	31/01/2023
					CT2750	Implantable cardiac electrical stimulators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	31/01/2023	31/01/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	31/01/2023	31/01/2023
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	31/01/2023	31/01/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	31/01/2023	31/01/2023
					CT2985	Active	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	31/01/2023	31/01/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	31/01/2023	31/01/2023
					CT3001	Implantable metal	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	31/01/2023	31/01/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	31/01/2023	31/01/2023
					CT3002	Implantable synthetic polymer	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	31/01/2023	31/01/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	31/01/2023	31/01/2023
					CT3006	Implantable inorganic material	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	31/01/2023	31/01/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	31/01/2023	31/01/2023
					CT314	Electrically powered	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	31/01/2023	31/01/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	31/01/2023	31/01/2023
					CT321	Long-term surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	31/01/2023	31/01/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	31/01/2023	31/01/2023
					CT449	Generators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	31/01/2023	31/01/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	31/01/2023	31/01/2023
					CT981	Single use	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	31/01/2023	31/01/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	31/01/2023	31/01/2023
					CT983	Surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	31/01/2023	31/01/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	31/01/2023	31/01/2023
12316	2023	All other reportable incidents	47265	Dual-chamber implantable pacemaker, rate-responsive	CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A0908	Incorrect, Inadequate or Imprecise Result or Readings		
							F26	No Health Consequences or Impact	A0908	Incorrect, Inadequate or Imprecise Result or Readings		
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A0908	Incorrect, Inadequate or Imprecise Result or Readings		
							F26	No Health Consequences or Impact	A0908	Incorrect, Inadequate or Imprecise Result or Readings		
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A0908	Incorrect, Inadequate or Imprecise Result or Readings		
							F26	No Health Consequences or Impact	A0908	Incorrect, Inadequate or Imprecise Result or Readings		
12317	2023		47265	Dual-chamber implantable pacemaker, rate-responsive	CT141	Electrodes/Leads and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	11/02/2023	11/02/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	11/02/2023	11/02/2023
					CT145	Cardiovascular devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	11/02/2023	11/02/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	11/02/2023	11/02/2023
					CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	11/02/2023	11/02/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	11/02/2023	11/02/2023
					CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	11/02/2023	11/02/2023

							F1901	Additional Surgery	A24	No Apparent Device Problem	11/02/2023	11/02/2023
					CT233	Surgical	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	11/02/2023	11/02/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	11/02/2023	11/02/2023
					CT2406	Implantable	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	11/02/2023	11/02/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	11/02/2023	11/02/2023
					CT2407	Active implantable	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	11/02/2023	11/02/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	11/02/2023	11/02/2023
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	11/02/2023	11/02/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	11/02/2023	11/02/2023
					CT2727	Active implant devices/ancillaries	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	11/02/2023	11/02/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	11/02/2023	11/02/2023
					CT2750	Implantable cardiac electrical stimulators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	11/02/2023	11/02/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	11/02/2023	11/02/2023
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	11/02/2023	11/02/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	11/02/2023	11/02/2023
					CT2985	Active	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	11/02/2023	11/02/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	11/02/2023	11/02/2023
					CT3001	Implantable metal	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	11/02/2023	11/02/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	11/02/2023	11/02/2023
					CT3002	Implantable synthetic polymer	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	11/02/2023	11/02/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	11/02/2023	11/02/2023
					CT3006	Implantable inorganic material	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	11/02/2023	11/02/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	11/02/2023	11/02/2023
					CT314	Electrically powered	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	11/02/2023	11/02/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	11/02/2023	11/02/2023
					CT321	Long-term surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	11/02/2023	11/02/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	11/02/2023	11/02/2023
					CT449	Generators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	11/02/2023	11/02/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	11/02/2023	11/02/2023
					CT981	Single use	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	11/02/2023	11/02/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	11/02/2023	11/02/2023
					CT983	Surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	11/02/2023	11/02/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	11/02/2023	11/02/2023
12318	2023		47265	Dual-chamber implantable pacemaker, rate-responsive	CT141	Electrodes/Leads and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	11/02/2023	11/02/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	11/02/2023	11/02/2023
					CT145	Cardiovascular devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	11/02/2023	11/02/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	11/02/2023	11/02/2023
					CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	11/02/2023	11/02/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	11/02/2023	11/02/2023
					CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	11/02/2023	11/02/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	11/02/2023	11/02/2023
					CT233	Surgical	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	11/02/2023	11/02/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	11/02/2023	11/02/2023
					CT2406	Implantable	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	11/02/2023	11/02/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	11/02/2023	11/02/2023
					CT2407	Active implantable	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	11/02/2023	11/02/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	11/02/2023	11/02/2023
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	11/02/2023	11/02/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	11/02/2023	11/02/2023
					CT2727	Active implant devices/ancillaries	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	11/02/2023	11/02/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	11/02/2023	11/02/2023
					CT2750	Implantable cardiac electrical stimulators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	11/02/2023	11/02/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	11/02/2023	11/02/2023
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	11/02/2023	11/02/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	11/02/2023	11/02/2023
					CT2985	Active	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	11/02/2023	11/02/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	11/02/2023	11/02/2023
					CT3001	Implantable metal	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	11/02/2023	11/02/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	11/02/2023	11/02/2023
					CT3002	Implantable synthetic polymer	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	11/02/2023	11/02/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	11/02/2023	11/02/2023
					CT3006	Implantable inorganic material	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	11/02/2023	11/02/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	11/02/2023	11/02/2023
					CT314	Electrically powered	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	11/02/2023	11/02/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	11/02/2023	11/02/2023
					CT321	Long-term surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	11/02/2023	11/02/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	11/02/2023	11/02/2023
					CT449	Generators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	11/02/2023	11/02/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	11/02/2023	11/02/2023
					CT981	Single use	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	11/02/2023	11/02/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	11/02/2023	11/02/2023
					CT983	Surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	11/02/2023	11/02/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	11/02/2023	11/02/2023
12319	2023		47265	Dual-chamber implantable pacemaker, rate-responsive	CT141	Electrodes/Leads and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	30/01/2023	30/01/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	30/01/2023	30/01/2023
					CT145	Cardiovascular devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	30/01/2023	30/01/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	30/01/2023	30/01/2023
					CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	30/01/2023	30/01/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	30/01/2023	30/01/2023
					CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	30/01/2023	30/01/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	30/01/2023	30/01/2023
					CT233	Surgical	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	30/01/2023	30/01/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	30/01/2023	30/01/2023
					CT2406	Implantable	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	30/01/2023	30/01/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	30/01/2023	30/01/2023
					CT2407	Active implantable	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	30/01/2023	30/01/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	30/01/2023	30/01/2023
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	30/01/2023	30/01/2023

						F1901	Additional Surgery	A24	No Apparent Device Problem	30/01/2023	30/01/2023
					CT2727	Active implant devices/ancillaries	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	30/01/2023	30/01/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	30/01/2023	30/01/2023
					CT2750	Implantable cardiac electrical stimulators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	30/01/2023	30/01/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	30/01/2023	30/01/2023
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	30/01/2023	30/01/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	30/01/2023	30/01/2023
					CT2985	Active	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	30/01/2023	30/01/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	30/01/2023	30/01/2023
					CT3001	Implantable metal	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	30/01/2023	30/01/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	30/01/2023	30/01/2023
					CT3002	Implantable synthetic polymer	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	30/01/2023	30/01/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	30/01/2023	30/01/2023
					CT3006	Implantable inorganic material	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	30/01/2023	30/01/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	30/01/2023	30/01/2023
					CT314	Electrically powered	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	30/01/2023	30/01/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	30/01/2023	30/01/2023
					CT321	Long-term surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	30/01/2023	30/01/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	30/01/2023	30/01/2023
					CT449	Generators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	30/01/2023	30/01/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	30/01/2023	30/01/2023
					CT981	Single use	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	30/01/2023	30/01/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	30/01/2023	30/01/2023
					CT983	Surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	30/01/2023	30/01/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	30/01/2023	30/01/2023
12320	2023		47265	Dual-chamber implantable pacemaker, rate-responsive	CT141	Electrodes/Leads and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	22/01/2023	22/01/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	22/01/2023	22/01/2023
					CT145	Cardiovascular devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	22/01/2023	22/01/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	22/01/2023	22/01/2023
					CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	22/01/2023	22/01/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	22/01/2023	22/01/2023
					CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	22/01/2023	22/01/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	22/01/2023	22/01/2023
					CT233	Surgical	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	22/01/2023	22/01/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	22/01/2023	22/01/2023
					CT2406	Implantable	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	22/01/2023	22/01/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	22/01/2023	22/01/2023
					CT2407	Active implantable	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	22/01/2023	22/01/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	22/01/2023	22/01/2023
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	22/01/2023	22/01/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	22/01/2023	22/01/2023
					CT2727	Active implant devices/ancillaries	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	22/01/2023	22/01/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	22/01/2023	22/01/2023
					CT2750	Implantable cardiac electrical stimulators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	22/01/2023	22/01/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	22/01/2023	22/01/2023
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	22/01/2023	22/01/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	22/01/2023	22/01/2023
					CT2985	Active	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	22/01/2023	22/01/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	22/01/2023	22/01/2023
					CT3001	Implantable metal	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	22/01/2023	22/01/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	22/01/2023	22/01/2023
					CT3002	Implantable synthetic polymer	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	22/01/2023	22/01/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	22/01/2023	22/01/2023
					CT3006	Implantable inorganic material	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	22/01/2023	22/01/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	22/01/2023	22/01/2023
					CT314	Electrically powered	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	22/01/2023	22/01/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	22/01/2023	22/01/2023
					CT321	Long-term surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	22/01/2023	22/01/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	22/01/2023	22/01/2023
					CT449	Generators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	22/01/2023	22/01/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	22/01/2023	22/01/2023
					CT981	Single use	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	22/01/2023	22/01/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	22/01/2023	22/01/2023
					CT983	Surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	22/01/2023	22/01/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	22/01/2023	22/01/2023
12321	2023	All other reportable incidents	35223	Endocardial/interventricular septal pacing lead	CT1684	Cardiac pacemakers and associated devices	E2103	Device Overstimulation of Tissue	A07	Electrical /Electronic Property Problem		
									A070102	High Capture Threshold		
									A070909	Oversensing		
									A072201	High Impedance		
									A090801	Signal Artifact/Noise		
							F23	Unexpected Medical Intervention	A07	Electrical /Electronic Property Problem		
									A070102	High Capture Threshold		
									A070909	Oversensing		
									A072201	High Impedance		
									A090801	Signal Artifact/Noise		
12322	2023	All other reportable incidents	36241	Pacemaker/defibrillator lead	CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A072201	High Impedance		
							F26	No Health Consequences or Impact	A072201	High Impedance		
12323	2023		47265	Dual-chamber implantable pacemaker, rate-responsive	CT141	Electrodes/Leads and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	02/02/2023	02/02/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	02/02/2023	02/02/2023
					CT145	Cardiovascular devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	02/02/2023	02/02/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	02/02/2023	02/02/2023
					CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	02/02/2023	02/02/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	02/02/2023	02/02/2023
					CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	02/02/2023	02/02/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	02/02/2023	02/02/2023
					CT233	Surgical	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	02/02/2023	02/02/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	02/02/2023	02/02/2023
					CT2406	Implantable	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	02/02/2023	02/02/2023

							F1901	Additional Surgery	A24	No Apparent Device Problem	02/02/2023	02/02/2023
					CT2407	Active implantable	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	02/02/2023	02/02/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	02/02/2023	02/02/2023
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	02/02/2023	02/02/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	02/02/2023	02/02/2023
					CT2727	Active implant devices/ancillaries	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	02/02/2023	02/02/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	02/02/2023	02/02/2023
					CT2750	Implantable cardiac electrical stimulators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	02/02/2023	02/02/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	02/02/2023	02/02/2023
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	02/02/2023	02/02/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	02/02/2023	02/02/2023
					CT2985	Active	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	02/02/2023	02/02/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	02/02/2023	02/02/2023
					CT3001	Implantable metal	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	02/02/2023	02/02/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	02/02/2023	02/02/2023
					CT3002	Implantable synthetic polymer	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	02/02/2023	02/02/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	02/02/2023	02/02/2023
					CT3006	Implantable inorganic material	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	02/02/2023	02/02/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	02/02/2023	02/02/2023
					CT314	Electrically powered	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	02/02/2023	02/02/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	02/02/2023	02/02/2023
					CT321	Long-term surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	02/02/2023	02/02/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	02/02/2023	02/02/2023
					CT449	Generators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	02/02/2023	02/02/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	02/02/2023	02/02/2023
					CT981	Single use	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	02/02/2023	02/02/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	02/02/2023	02/02/2023
					CT983	Surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	02/02/2023	02/02/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	02/02/2023	02/02/2023
12324	2023	All other reportable incidents	35852	Single-chamber implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
12325	2023		47265	Dual-chamber implantable pacemaker, rate-responsive	CT141	Electrodes/Leads and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	15/01/2023	15/01/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	15/01/2023	15/01/2023
					CT145	Cardiovascular devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	15/01/2023	15/01/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	15/01/2023	15/01/2023
					CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	15/01/2023	15/01/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	15/01/2023	15/01/2023
					CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	15/01/2023	15/01/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	15/01/2023	15/01/2023
					CT233	Surgical	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	15/01/2023	15/01/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	15/01/2023	15/01/2023
					CT2406	Implantable	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	15/01/2023	15/01/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	15/01/2023	15/01/2023
					CT2407	Active implantable	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	15/01/2023	15/01/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	15/01/2023	15/01/2023
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	15/01/2023	15/01/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	15/01/2023	15/01/2023
					CT2727	Active implant devices/ancillaries	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	15/01/2023	15/01/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	15/01/2023	15/01/2023
					CT2750	Implantable cardiac electrical stimulators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	15/01/2023	15/01/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	15/01/2023	15/01/2023
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	15/01/2023	15/01/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	15/01/2023	15/01/2023
					CT2985	Active	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	15/01/2023	15/01/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	15/01/2023	15/01/2023
					CT3001	Implantable metal	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	15/01/2023	15/01/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	15/01/2023	15/01/2023
					CT3002	Implantable synthetic polymer	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	15/01/2023	15/01/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	15/01/2023	15/01/2023
					CT3006	Implantable inorganic material	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	15/01/2023	15/01/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	15/01/2023	15/01/2023
					CT314	Electrically powered	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	15/01/2023	15/01/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	15/01/2023	15/01/2023
					CT321	Long-term surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	15/01/2023	15/01/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	15/01/2023	15/01/2023
					CT449	Generators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	15/01/2023	15/01/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	15/01/2023	15/01/2023
					CT981	Single use	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	15/01/2023	15/01/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	15/01/2023	15/01/2023
					CT983	Surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	15/01/2023	15/01/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	15/01/2023	15/01/2023
12326	2023	All other reportable incidents	47265	Dual-chamber implantable pacemaker, rate-responsive	CT141	Electrodes/Leads and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	26/01/2023	26/01/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	26/01/2023	26/01/2023
					CT145	Cardiovascular devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	26/01/2023	26/01/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	26/01/2023	26/01/2023
					CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	26/01/2023	26/01/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	26/01/2023	26/01/2023
					CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	26/01/2023	26/01/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	26/01/2023	26/01/2023
					CT233	Surgical	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	26/01/2023	26/01/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	26/01/2023	26/01/2023
					CT2406	Implantable	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	26/01/2023	26/01/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	26/01/2023	26/01/2023
					CT2407	Active implantable	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	26/01/2023	26/01/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	26/01/2023	26/01/2023
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	26/01/2023	26/01/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	26/01/2023	26/01/2023
					CT2727	Active implant devices/ancillaries	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	26/01/2023	26/01/2023

							F1901	Additional Surgery	A24	No Apparent Device Problem	26/01/2023	26/01/2023
					CT2750	Implantable cardiac electrical stimulators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	26/01/2023	26/01/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	26/01/2023	26/01/2023
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	26/01/2023	26/01/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	26/01/2023	26/01/2023
					CT2985	Active	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	26/01/2023	26/01/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	26/01/2023	26/01/2023
					CT3001	Implantable metal	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	26/01/2023	26/01/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	26/01/2023	26/01/2023
					CT3002	Implantable synthetic polymer	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	26/01/2023	26/01/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	26/01/2023	26/01/2023
					CT3006	Implantable inorganic material	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	26/01/2023	26/01/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	26/01/2023	26/01/2023
					CT314	Electrically powered	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	26/01/2023	26/01/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	26/01/2023	26/01/2023
					CT321	Long-term surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	26/01/2023	26/01/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	26/01/2023	26/01/2023
					CT449	Generators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	26/01/2023	26/01/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	26/01/2023	26/01/2023
					CT981	Single use	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	26/01/2023	26/01/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	26/01/2023	26/01/2023
					CT983	Surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	26/01/2023	26/01/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	26/01/2023	26/01/2023
12327	2023		47265	Dual-chamber implantable pacemaker, rate-responsive	CT141	Electrodes/Leads and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	11/02/2023	11/02/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	11/02/2023	11/02/2023
					CT145	Cardiovascular devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	11/02/2023	11/02/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	11/02/2023	11/02/2023
					CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	11/02/2023	11/02/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	11/02/2023	11/02/2023
					CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	11/02/2023	11/02/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	11/02/2023	11/02/2023
					CT233	Surgical	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	11/02/2023	11/02/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	11/02/2023	11/02/2023
					CT2406	Implantable	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	11/02/2023	11/02/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	11/02/2023	11/02/2023
					CT2407	Active implantable	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	11/02/2023	11/02/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	11/02/2023	11/02/2023
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	11/02/2023	11/02/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	11/02/2023	11/02/2023
					CT2727	Active implant devices/ancillaries	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	11/02/2023	11/02/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	11/02/2023	11/02/2023
					CT2750	Implantable cardiac electrical stimulators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	11/02/2023	11/02/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	11/02/2023	11/02/2023
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	11/02/2023	11/02/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	11/02/2023	11/02/2023
					CT2985	Active	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	11/02/2023	11/02/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	11/02/2023	11/02/2023
					CT3001	Implantable metal	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	11/02/2023	11/02/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	11/02/2023	11/02/2023
					CT3002	Implantable synthetic polymer	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	11/02/2023	11/02/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	11/02/2023	11/02/2023
					CT3006	Implantable inorganic material	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	11/02/2023	11/02/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	11/02/2023	11/02/2023
					CT314	Electrically powered	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	11/02/2023	11/02/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	11/02/2023	11/02/2023
					CT321	Long-term surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	11/02/2023	11/02/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	11/02/2023	11/02/2023
					CT449	Generators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	11/02/2023	11/02/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	11/02/2023	11/02/2023
					CT981	Single use	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	11/02/2023	11/02/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	11/02/2023	11/02/2023
					CT983	Surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	11/02/2023	11/02/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	11/02/2023	11/02/2023
12328	2023	All other reportable incidents	37265	Dual-chamber implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	12/09/2022	12/09/2022
									A160601	Premature Elective Replacement Indicator	12/09/2022	12/09/2022
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	12/09/2022	12/09/2022
									A160601	Premature Elective Replacement Indicator	12/09/2022	12/09/2022
							F1903	Device Explantation	A070504	Premature Discharge of Battery	12/09/2022	12/09/2022
									A160601	Premature Elective Replacement Indicator	12/09/2022	12/09/2022
12329	2023		47265	Dual-chamber implantable pacemaker, rate-responsive	CT141	Electrodes/Leads and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	17/01/2023	17/01/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	17/01/2023	17/01/2023
					CT145	Cardiovascular devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	17/01/2023	17/01/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	17/01/2023	17/01/2023
					CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	17/01/2023	17/01/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	17/01/2023	17/01/2023
					CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	17/01/2023	17/01/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	17/01/2023	17/01/2023
					CT233	Surgical	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	17/01/2023	17/01/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	17/01/2023	17/01/2023
					CT2406	Implantable	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	17/01/2023	17/01/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	17/01/2023	17/01/2023
					CT2407	Active implantable	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	17/01/2023	17/01/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	17/01/2023	17/01/2023
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	17/01/2023	17/01/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	17/01/2023	17/01/2023
					CT2727	Active implant devices/ancillaries	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	17/01/2023	17/01/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	17/01/2023	17/01/2023
					CT2750	Implantable cardiac electrical stimulators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	17/01/2023	17/01/2023

							F1901	Additional Surgery	A24	No Apparent Device Problem	17/01/2023	17/01/2023
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	17/01/2023	17/01/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	17/01/2023	17/01/2023
					CT2985	Active	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	17/01/2023	17/01/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	17/01/2023	17/01/2023
					CT3001	Implantable metal	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	17/01/2023	17/01/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	17/01/2023	17/01/2023
					CT3002	Implantable synthetic polymer	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	17/01/2023	17/01/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	17/01/2023	17/01/2023
					CT3006	Implantable inorganic material	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	17/01/2023	17/01/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	17/01/2023	17/01/2023
					CT314	Electrically powered	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	17/01/2023	17/01/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	17/01/2023	17/01/2023
					CT321	Long-term surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	17/01/2023	17/01/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	17/01/2023	17/01/2023
					CT449	Generators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	17/01/2023	17/01/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	17/01/2023	17/01/2023
					CT981	Single use	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	17/01/2023	17/01/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	17/01/2023	17/01/2023
					CT983	Surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	17/01/2023	17/01/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	17/01/2023	17/01/2023
12330	2023		47265	Dual-chamber implantable pacemaker, rate-responsive	CT141	Electrodes/Leads and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	11/02/2023	11/02/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	11/02/2023	11/02/2023
					CT145	Cardiovascular devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	11/02/2023	11/02/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	11/02/2023	11/02/2023
					CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	11/02/2023	11/02/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	11/02/2023	11/02/2023
					CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	11/02/2023	11/02/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	11/02/2023	11/02/2023
					CT233	Surgical	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	11/02/2023	11/02/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	11/02/2023	11/02/2023
					CT2406	Implantable	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	11/02/2023	11/02/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	11/02/2023	11/02/2023
					CT2407	Active implantable	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	11/02/2023	11/02/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	11/02/2023	11/02/2023
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	11/02/2023	11/02/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	11/02/2023	11/02/2023
					CT2727	Active implant devices/ancillaries	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	11/02/2023	11/02/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	11/02/2023	11/02/2023
					CT2750	Implantable cardiac electrical stimulators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	11/02/2023	11/02/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	11/02/2023	11/02/2023
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	11/02/2023	11/02/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	11/02/2023	11/02/2023
					CT2985	Active	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	11/02/2023	11/02/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	11/02/2023	11/02/2023
					CT3001	Implantable metal	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	11/02/2023	11/02/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	11/02/2023	11/02/2023
					CT3002	Implantable synthetic polymer	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	11/02/2023	11/02/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	11/02/2023	11/02/2023
					CT3006	Implantable inorganic material	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	11/02/2023	11/02/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	11/02/2023	11/02/2023
					CT314	Electrically powered	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	11/02/2023	11/02/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	11/02/2023	11/02/2023
					CT321	Long-term surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	11/02/2023	11/02/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	11/02/2023	11/02/2023
					CT449	Generators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	11/02/2023	11/02/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	11/02/2023	11/02/2023
					CT981	Single use	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	11/02/2023	11/02/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	11/02/2023	11/02/2023
					CT983	Surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	11/02/2023	11/02/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	11/02/2023	11/02/2023
12331	2023		47265	Dual-chamber implantable pacemaker, rate-responsive	CT141	Electrodes/Leads and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	31/01/2023	31/01/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	31/01/2023	31/01/2023
					CT145	Cardiovascular devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	31/01/2023	31/01/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	31/01/2023	31/01/2023
					CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	31/01/2023	31/01/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	31/01/2023	31/01/2023
					CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	31/01/2023	31/01/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	31/01/2023	31/01/2023
					CT233	Surgical	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	31/01/2023	31/01/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	31/01/2023	31/01/2023
					CT2406	Implantable	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	31/01/2023	31/01/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	31/01/2023	31/01/2023
					CT2407	Active implantable	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	31/01/2023	31/01/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	31/01/2023	31/01/2023
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	31/01/2023	31/01/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	31/01/2023	31/01/2023
					CT2727	Active implant devices/ancillaries	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	31/01/2023	31/01/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	31/01/2023	31/01/2023
					CT2750	Implantable cardiac electrical stimulators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	31/01/2023	31/01/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	31/01/2023	31/01/2023
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	31/01/2023	31/01/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	31/01/2023	31/01/2023
					CT2985	Active	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	31/01/2023	31/01/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	31/01/2023	31/01/2023
					CT3001	Implantable metal	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	31/01/2023	31/01/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	31/01/2023	31/01/2023
					CT3002	Implantable synthetic polymer	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	31/01/2023	31/01/2023

						F1901	Additional Surgery	A24	No Apparent Device Problem	31/01/2023	31/01/2023
				CT3006	Implantable inorganic material	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	31/01/2023	31/01/2023
						F1901	Additional Surgery	A24	No Apparent Device Problem	31/01/2023	31/01/2023
				CT314	Electrically powered	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	31/01/2023	31/01/2023
						F1901	Additional Surgery	A24	No Apparent Device Problem	31/01/2023	31/01/2023
				CT321	Long-term surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	31/01/2023	31/01/2023
						F1901	Additional Surgery	A24	No Apparent Device Problem	31/01/2023	31/01/2023
				CT449	Generators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	31/01/2023	31/01/2023
						F1901	Additional Surgery	A24	No Apparent Device Problem	31/01/2023	31/01/2023
				CT981	Single use	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	31/01/2023	31/01/2023
						F1901	Additional Surgery	A24	No Apparent Device Problem	31/01/2023	31/01/2023
				CT983	Surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	31/01/2023	31/01/2023
						F1901	Additional Surgery	A24	No Apparent Device Problem	31/01/2023	31/01/2023
12332	2023	All other reportable incidents	35852	Single-chamber implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
12333	2023	All other reportable incidents	35223	Endocardial/interventricular septal pacing lead	CT1684	Cardiac pacemakers and associated devices	E0604	Cardiac Perforation	A051201	Device Dislodged or Dislocated		
									A0701	Capturing Problem		
									A072201	High Impedance		
							F22	Unexpected Diagnostic Intervention	A051201	Device Dislodged or Dislocated		
									A0701	Capturing Problem		
									A072201	High Impedance		
12334	2023	All other reportable incidents	47263	Cardiac resynchronization therapy implantable pacemaker	CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A0203	Defective Device	16/09/2021	16/09/2021
							F24	Insufficient Health Impact Information	A0203	Defective Device	16/09/2021	16/09/2021
					CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A0203	Defective Device	16/09/2021	16/09/2021
							F24	Insufficient Health Impact Information	A0203	Defective Device	16/09/2021	16/09/2021
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A0203	Defective Device	16/09/2021	16/09/2021
							F24	Insufficient Health Impact Information	A0203	Defective Device	16/09/2021	16/09/2021
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A0203	Defective Device	16/09/2021	16/09/2021
							F24	Insufficient Health Impact Information	A0203	Defective Device	16/09/2021	16/09/2021
12335	2023	All other reportable incidents	17882	Physiologic-monitoring defibrillation system	CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070804	Unintended Power Up		
							F27	Problem Identified During Non-Clinical Procedure	A070804	Unintended Power Up		
12336	2023		47265	Dual-chamber implantable pacemaker, rate-responsive	CT141	Electrodes/Leads and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	03/02/2023	03/02/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	03/02/2023	03/02/2023
					CT145	Cardiovascular devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	03/02/2023	03/02/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	03/02/2023	03/02/2023
					CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	03/02/2023	03/02/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	03/02/2023	03/02/2023
					CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	03/02/2023	03/02/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	03/02/2023	03/02/2023
					CT233	Surgical	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	03/02/2023	03/02/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	03/02/2023	03/02/2023
					CT2406	Implantable	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	03/02/2023	03/02/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	03/02/2023	03/02/2023
					CT2407	Active implantable	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	03/02/2023	03/02/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	03/02/2023	03/02/2023
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	03/02/2023	03/02/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	03/02/2023	03/02/2023
					CT2727	Active implant devices/ancillaries	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	03/02/2023	03/02/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	03/02/2023	03/02/2023
					CT2750	Implantable cardiac electrical stimulators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	03/02/2023	03/02/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	03/02/2023	03/02/2023
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	03/02/2023	03/02/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	03/02/2023	03/02/2023
					CT2985	Active	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	03/02/2023	03/02/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	03/02/2023	03/02/2023
					CT3001	Implantable metal	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	03/02/2023	03/02/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	03/02/2023	03/02/2023
					CT3002	Implantable synthetic polymer	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	03/02/2023	03/02/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	03/02/2023	03/02/2023
					CT3006	Implantable inorganic material	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	03/02/2023	03/02/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	03/02/2023	03/02/2023
					CT314	Electrically powered	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	03/02/2023	03/02/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	03/02/2023	03/02/2023
					CT321	Long-term surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	03/02/2023	03/02/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	03/02/2023	03/02/2023
					CT449	Generators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	03/02/2023	03/02/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	03/02/2023	03/02/2023
					CT981	Single use	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	03/02/2023	03/02/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	03/02/2023	03/02/2023
					CT983	Surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	03/02/2023	03/02/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	03/02/2023	03/02/2023
12337	2023		47265	Dual-chamber implantable pacemaker, rate-responsive	CT141	Electrodes/Leads and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	03/02/2023	03/02/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	03/02/2023	03/02/2023
					CT145	Cardiovascular devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	03/02/2023	03/02/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	03/02/2023	03/02/2023
					CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	03/02/2023	03/02/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	03/02/2023	03/02/2023
					CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	03/02/2023	03/02/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	03/02/2023	03/02/2023
					CT233	Surgical	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	03/02/2023	03/02/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	03/02/2023	03/02/2023
					CT2406	Implantable	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	03/02/2023	03/02/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	03/02/2023	03/02/2023
					CT2407	Active implantable	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	03/02/2023	03/02/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	03/02/2023	03/02/2023
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	03/02/2023	03/02/2023

						F1901	Additional Surgery	A24	No Apparent Device Problem	03/02/2023	03/02/2023
					CT2727	Active implant devices/ancillaries	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	03/02/2023	03/02/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	03/02/2023	03/02/2023
					CT2750	Implantable cardiac electrical stimulators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	03/02/2023	03/02/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	03/02/2023	03/02/2023
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	03/02/2023	03/02/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	03/02/2023	03/02/2023
					CT2985	Active	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	03/02/2023	03/02/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	03/02/2023	03/02/2023
					CT3001	Implantable metal	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	03/02/2023	03/02/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	03/02/2023	03/02/2023
					CT3002	Implantable synthetic polymer	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	03/02/2023	03/02/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	03/02/2023	03/02/2023
					CT3006	Implantable inorganic material	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	03/02/2023	03/02/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	03/02/2023	03/02/2023
					CT314	Electrically powered	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	03/02/2023	03/02/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	03/02/2023	03/02/2023
					CT321	Long-term surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	03/02/2023	03/02/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	03/02/2023	03/02/2023
					CT449	Generators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	03/02/2023	03/02/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	03/02/2023	03/02/2023
					CT981	Single use	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	03/02/2023	03/02/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	03/02/2023	03/02/2023
					CT983	Surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	03/02/2023	03/02/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	03/02/2023	03/02/2023
12338	2023		47265	Dual-chamber implantable pacemaker, rate-responsive	CT141	Electrodes/Leads and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	11/02/2023	11/02/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	11/02/2023	11/02/2023
					CT145	Cardiovascular devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	11/02/2023	11/02/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	11/02/2023	11/02/2023
					CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	11/02/2023	11/02/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	11/02/2023	11/02/2023
					CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	11/02/2023	11/02/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	11/02/2023	11/02/2023
					CT233	Surgical	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	11/02/2023	11/02/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	11/02/2023	11/02/2023
					CT2406	Implantable	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	11/02/2023	11/02/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	11/02/2023	11/02/2023
					CT2407	Active implantable	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	11/02/2023	11/02/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	11/02/2023	11/02/2023
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	11/02/2023	11/02/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	11/02/2023	11/02/2023
					CT2727	Active implant devices/ancillaries	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	11/02/2023	11/02/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	11/02/2023	11/02/2023
					CT2750	Implantable cardiac electrical stimulators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	11/02/2023	11/02/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	11/02/2023	11/02/2023
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	11/02/2023	11/02/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	11/02/2023	11/02/2023
					CT2985	Active	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	11/02/2023	11/02/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	11/02/2023	11/02/2023
					CT3001	Implantable metal	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	11/02/2023	11/02/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	11/02/2023	11/02/2023
					CT3002	Implantable synthetic polymer	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	11/02/2023	11/02/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	11/02/2023	11/02/2023
					CT3006	Implantable inorganic material	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	11/02/2023	11/02/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	11/02/2023	11/02/2023
					CT314	Electrically powered	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	11/02/2023	11/02/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	11/02/2023	11/02/2023
					CT321	Long-term surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	11/02/2023	11/02/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	11/02/2023	11/02/2023
					CT449	Generators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	11/02/2023	11/02/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	11/02/2023	11/02/2023
					CT981	Single use	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	11/02/2023	11/02/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	11/02/2023	11/02/2023
					CT983	Surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	11/02/2023	11/02/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	11/02/2023	11/02/2023
12339	2023		47265	Dual-chamber implantable pacemaker, rate-responsive	CT141	Electrodes/Leads and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	11/02/2023	11/02/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	11/02/2023	11/02/2023
					CT145	Cardiovascular devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	11/02/2023	11/02/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	11/02/2023	11/02/2023
					CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	11/02/2023	11/02/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	11/02/2023	11/02/2023
					CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	11/02/2023	11/02/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	11/02/2023	11/02/2023
					CT233	Surgical	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	11/02/2023	11/02/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	11/02/2023	11/02/2023
					CT2406	Implantable	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	11/02/2023	11/02/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	11/02/2023	11/02/2023
					CT2407	Active implantable	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	11/02/2023	11/02/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	11/02/2023	11/02/2023
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	11/02/2023	11/02/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	11/02/2023	11/02/2023
					CT2727	Active implant devices/ancillaries	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	11/02/2023	11/02/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	11/02/2023	11/02/2023
					CT2750	Implantable cardiac electrical stimulators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	11/02/2023	11/02/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	11/02/2023	11/02/2023
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	11/02/2023	11/02/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	11/02/2023	11/02/2023
					CT2985	Active	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	11/02/2023	11/02/2023

							F1901	Additional Surgery	A24	No Apparent Device Problem	11/02/2023	11/02/2023
					CT3001	Implantable metal	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	11/02/2023	11/02/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	11/02/2023	11/02/2023
					CT3002	Implantable synthetic polymer	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	11/02/2023	11/02/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	11/02/2023	11/02/2023
					CT3006	Implantable inorganic material	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	11/02/2023	11/02/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	11/02/2023	11/02/2023
					CT314	Electrically powered	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	11/02/2023	11/02/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	11/02/2023	11/02/2023
					CT321	Long-term surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	11/02/2023	11/02/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	11/02/2023	11/02/2023
					CT449	Generators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	11/02/2023	11/02/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	11/02/2023	11/02/2023
					CT981	Single use	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	11/02/2023	11/02/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	11/02/2023	11/02/2023
					CT983	Surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	11/02/2023	11/02/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	11/02/2023	11/02/2023
12340	2023		47265	Dual-chamber implantable pacemaker, rate-responsive	CT141	Electrodes/Leads and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	11/02/2023	11/02/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	11/02/2023	11/02/2023
					CT145	Cardiovascular devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	11/02/2023	11/02/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	11/02/2023	11/02/2023
					CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	11/02/2023	11/02/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	11/02/2023	11/02/2023
					CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	11/02/2023	11/02/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	11/02/2023	11/02/2023
					CT233	Surgical	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	11/02/2023	11/02/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	11/02/2023	11/02/2023
					CT2406	Implantable	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	11/02/2023	11/02/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	11/02/2023	11/02/2023
					CT2407	Active implantable	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	11/02/2023	11/02/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	11/02/2023	11/02/2023
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	11/02/2023	11/02/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	11/02/2023	11/02/2023
					CT2727	Active implant devices/ancillaries	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	11/02/2023	11/02/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	11/02/2023	11/02/2023
					CT2750	Implantable cardiac electrical stimulators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	11/02/2023	11/02/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	11/02/2023	11/02/2023
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	11/02/2023	11/02/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	11/02/2023	11/02/2023
					CT2985	Active	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	11/02/2023	11/02/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	11/02/2023	11/02/2023
					CT3001	Implantable metal	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	11/02/2023	11/02/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	11/02/2023	11/02/2023
					CT3002	Implantable synthetic polymer	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	11/02/2023	11/02/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	11/02/2023	11/02/2023
					CT3006	Implantable inorganic material	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	11/02/2023	11/02/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	11/02/2023	11/02/2023
					CT314	Electrically powered	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	11/02/2023	11/02/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	11/02/2023	11/02/2023
					CT321	Long-term surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	11/02/2023	11/02/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	11/02/2023	11/02/2023
					CT449	Generators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	11/02/2023	11/02/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	11/02/2023	11/02/2023
					CT981	Single use	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	11/02/2023	11/02/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	11/02/2023	11/02/2023
					CT983	Surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	11/02/2023	11/02/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	11/02/2023	11/02/2023
12341	2023	All other reportable incidents	47270	Cardiac resynchronization therapy implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E011903	Syncope/Fainting	A070909	Oversensing		
							F23	Unexpected Medical Intervention	A070909	Oversensing		
12342	2023		47265	Dual-chamber implantable pacemaker, rate-responsive	CT141	Electrodes/Leads and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	08/01/2023	08/01/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	08/01/2023	08/01/2023
					CT145	Cardiovascular devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	08/01/2023	08/01/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	08/01/2023	08/01/2023
					CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	08/01/2023	08/01/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	08/01/2023	08/01/2023
					CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	08/01/2023	08/01/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	08/01/2023	08/01/2023
					CT233	Surgical	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	08/01/2023	08/01/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	08/01/2023	08/01/2023
					CT2406	Implantable	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	08/01/2023	08/01/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	08/01/2023	08/01/2023
					CT2407	Active implantable	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	08/01/2023	08/01/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	08/01/2023	08/01/2023
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	08/01/2023	08/01/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	08/01/2023	08/01/2023
					CT2727	Active implant devices/ancillaries	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	08/01/2023	08/01/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	08/01/2023	08/01/2023
					CT2750	Implantable cardiac electrical stimulators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	08/01/2023	08/01/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	08/01/2023	08/01/2023
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	08/01/2023	08/01/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	08/01/2023	08/01/2023
					CT2985	Active	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	08/01/2023	08/01/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	08/01/2023	08/01/2023
					CT3001	Implantable metal	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	08/01/2023	08/01/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	08/01/2023	08/01/2023
					CT3002	Implantable synthetic polymer	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	08/01/2023	08/01/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	08/01/2023	08/01/2023
					CT3006	Implantable inorganic material	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	08/01/2023	08/01/2023

							F1901	Additional Surgery	A24	No Apparent Device Problem	08/01/2023	08/01/2023
					CT314	Electrically powered	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	08/01/2023	08/01/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	08/01/2023	08/01/2023
					CT321	Long-term surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	08/01/2023	08/01/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	08/01/2023	08/01/2023
					CT449	Generators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	08/01/2023	08/01/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	08/01/2023	08/01/2023
					CT981	Single use	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	08/01/2023	08/01/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	08/01/2023	08/01/2023
					CT983	Surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	08/01/2023	08/01/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	08/01/2023	08/01/2023
12343	2023		47265	Dual-chamber implantable pacemaker, rate-responsive	CT141	Electrodes/Leads and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	28/01/2023	28/01/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	28/01/2023	28/01/2023
					CT145	Cardiovascular devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	28/01/2023	28/01/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	28/01/2023	28/01/2023
					CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	28/01/2023	28/01/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	28/01/2023	28/01/2023
					CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	28/01/2023	28/01/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	28/01/2023	28/01/2023
					CT233	Surgical	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	28/01/2023	28/01/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	28/01/2023	28/01/2023
					CT2406	Implantable	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	28/01/2023	28/01/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	28/01/2023	28/01/2023
					CT2407	Active implantable	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	28/01/2023	28/01/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	28/01/2023	28/01/2023
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	28/01/2023	28/01/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	28/01/2023	28/01/2023
					CT2727	Active implant devices/ancillaries	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	28/01/2023	28/01/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	28/01/2023	28/01/2023
					CT2750	Implantable cardiac electrical stimulators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	28/01/2023	28/01/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	28/01/2023	28/01/2023
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	28/01/2023	28/01/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	28/01/2023	28/01/2023
					CT2985	Active	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	28/01/2023	28/01/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	28/01/2023	28/01/2023
					CT3001	Implantable metal	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	28/01/2023	28/01/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	28/01/2023	28/01/2023
					CT3002	Implantable synthetic polymer	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	28/01/2023	28/01/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	28/01/2023	28/01/2023
					CT3006	Implantable inorganic material	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	28/01/2023	28/01/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	28/01/2023	28/01/2023
					CT314	Electrically powered	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	28/01/2023	28/01/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	28/01/2023	28/01/2023
					CT321	Long-term surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	28/01/2023	28/01/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	28/01/2023	28/01/2023
					CT449	Generators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	28/01/2023	28/01/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	28/01/2023	28/01/2023
					CT981	Single use	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	28/01/2023	28/01/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	28/01/2023	28/01/2023
					CT983	Surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	28/01/2023	28/01/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	28/01/2023	28/01/2023
12344	2023	All other reportable incidents	47265	Dual-chamber implantable pacemaker, rate-responsive	CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
12345	2023	All other reportable incidents	35852	Single-chamber implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
									A0908	Incorrect, Inadequate or Imprecise Result or Readings		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
									A0908	Incorrect, Inadequate or Imprecise Result or Readings		
12346	2023	All other reportable incidents	47265	Dual-chamber implantable pacemaker, rate-responsive	CT141	Electrodes/Leads and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A071204	No Pacing	10/02/2023	10/02/2023
							F1901	Additional Surgery	A071204	No Pacing	10/02/2023	10/02/2023
					CT145	Cardiovascular devices	E2403	No Clinical Signs, Symptoms or Conditions	A071204	No Pacing	10/02/2023	10/02/2023
							F1901	Additional Surgery	A071204	No Pacing	10/02/2023	10/02/2023
					CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A071204	No Pacing	10/02/2023	10/02/2023
							F1901	Additional Surgery	A071204	No Pacing	10/02/2023	10/02/2023
					CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A071204	No Pacing	10/02/2023	10/02/2023
							F1901	Additional Surgery	A071204	No Pacing	10/02/2023	10/02/2023
					CT233	Surgical	E2403	No Clinical Signs, Symptoms or Conditions	A071204	No Pacing	10/02/2023	10/02/2023
							F1901	Additional Surgery	A071204	No Pacing	10/02/2023	10/02/2023
					CT2406	Implantable	E2403	No Clinical Signs, Symptoms or Conditions	A071204	No Pacing	10/02/2023	10/02/2023
							F1901	Additional Surgery	A071204	No Pacing	10/02/2023	10/02/2023
					CT2407	Active implantable	E2403	No Clinical Signs, Symptoms or Conditions	A071204	No Pacing	10/02/2023	10/02/2023
							F1901	Additional Surgery	A071204	No Pacing	10/02/2023	10/02/2023
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A071204	No Pacing	10/02/2023	10/02/2023
							F1901	Additional Surgery	A071204	No Pacing	10/02/2023	10/02/2023
					CT2727	Active implant devices/ancillaries	E2403	No Clinical Signs, Symptoms or Conditions	A071204	No Pacing	10/02/2023	10/02/2023
							F1901	Additional Surgery	A071204	No Pacing	10/02/2023	10/02/2023
					CT2750	Implantable cardiac electrical stimulators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A071204	No Pacing	10/02/2023	10/02/2023

						F1901	Additional Surgery	A071204	No Pacing	10/02/2023	10/02/2023
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A071204	No Pacing	10/02/2023	10/02/2023
							F1901	Additional Surgery	A071204	No Pacing	10/02/2023	10/02/2023
					CT2985	Active	E2403	No Clinical Signs, Symptoms or Conditions	A071204	No Pacing	10/02/2023	10/02/2023
							F1901	Additional Surgery	A071204	No Pacing	10/02/2023	10/02/2023
					CT3001	Implantable metal	E2403	No Clinical Signs, Symptoms or Conditions	A071204	No Pacing	10/02/2023	10/02/2023
							F1901	Additional Surgery	A071204	No Pacing	10/02/2023	10/02/2023
					CT3002	Implantable synthetic polymer	E2403	No Clinical Signs, Symptoms or Conditions	A071204	No Pacing	10/02/2023	10/02/2023
							F1901	Additional Surgery	A071204	No Pacing	10/02/2023	10/02/2023
					CT3006	Implantable inorganic material	E2403	No Clinical Signs, Symptoms or Conditions	A071204	No Pacing	10/02/2023	10/02/2023
							F1901	Additional Surgery	A071204	No Pacing	10/02/2023	10/02/2023
					CT314	Electrically powered	E2403	No Clinical Signs, Symptoms or Conditions	A071204	No Pacing	10/02/2023	10/02/2023
							F1901	Additional Surgery	A071204	No Pacing	10/02/2023	10/02/2023
					CT321	Long-term surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A071204	No Pacing	10/02/2023	10/02/2023
							F1901	Additional Surgery	A071204	No Pacing	10/02/2023	10/02/2023
					CT449	Generators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A071204	No Pacing	10/02/2023	10/02/2023
							F1901	Additional Surgery	A071204	No Pacing	10/02/2023	10/02/2023
					CT981	Single use	E2403	No Clinical Signs, Symptoms or Conditions	A071204	No Pacing	10/02/2023	10/02/2023
							F1901	Additional Surgery	A071204	No Pacing	10/02/2023	10/02/2023
					CT983	Surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A071204	No Pacing	10/02/2023	10/02/2023
							F1901	Additional Surgery	A071204	No Pacing	10/02/2023	10/02/2023
12347	2023	All other reportable incidents	37265	Dual-chamber implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E2104	Electric Shock	A071302	Inappropriate/Inadequate Shock/Stimulation		
									A072201	High Impedance		
									A0908	Incorrect, Inadequate or Imprecise Result or Readings		
									A1301	Failure to Read Input Signal		
							F08	Hospitalization or Prolonged Hospitalization	A071302	Inappropriate/Inadequate Shock/Stimulation		
									A072201	High Impedance		
									A0908	Incorrect, Inadequate or Imprecise Result or Readings		
									A1301	Failure to Read Input Signal		
							F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A071302	Inappropriate/Inadequate Shock/Stimulation		
									A072201	High Impedance		
									A0908	Incorrect, Inadequate or Imprecise Result or Readings		
									A1301	Failure to Read Input Signal		
12348	2023	All other reportable incidents	47265	Dual-chamber implantable pacemaker, rate-responsive	CT141	Electrodes/Leads and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	11/01/2023	11/01/2023
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	11/01/2023	11/01/2023
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	11/01/2023	11/01/2023
					CT145	Cardiovascular devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	11/01/2023	11/01/2023
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	11/01/2023	11/01/2023
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	11/01/2023	11/01/2023
					CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	11/01/2023	11/01/2023
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	11/01/2023	11/01/2023
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	11/01/2023	11/01/2023
					CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	11/01/2023	11/01/2023
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	11/01/2023	11/01/2023
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	11/01/2023	11/01/2023
					CT233	Surgical	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	11/01/2023	11/01/2023
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	11/01/2023	11/01/2023
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	11/01/2023	11/01/2023
					CT2406	Implantable	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	11/01/2023	11/01/2023
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	11/01/2023	11/01/2023
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	11/01/2023	11/01/2023
					CT2407	Active implantable	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	11/01/2023	11/01/2023
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	11/01/2023	11/01/2023
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	11/01/2023	11/01/2023
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	11/01/2023	11/01/2023
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	11/01/2023	11/01/2023
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	11/01/2023	11/01/2023
					CT2727	Active implant devices/ancillaries	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	11/01/2023	11/01/2023
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	11/01/2023	11/01/2023
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	11/01/2023	11/01/2023
					CT2750	Implantable cardiac electrical stimulators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	11/01/2023	11/01/2023
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	11/01/2023	11/01/2023
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	11/01/2023	11/01/2023
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	11/01/2023	11/01/2023
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	11/01/2023	11/01/2023
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	11/01/2023	11/01/2023
					CT2985	Active	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	11/01/2023	11/01/2023
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	11/01/2023	11/01/2023
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	11/01/2023	11/01/2023
					CT3001	Implantable metal	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	11/01/2023	11/01/2023
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	11/01/2023	11/01/2023
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	11/01/2023	11/01/2023
					CT3002	Implantable synthetic polymer	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	11/01/2023	11/01/2023
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	11/01/2023	11/01/2023
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	11/01/2023	11/01/2023
					CT3006	Implantable inorganic material	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	11/01/2023	11/01/2023
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	11/01/2023	11/01/2023
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	11/01/2023	11/01/2023
					CT314	Electrically powered	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	11/01/2023	11/01/2023
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	11/01/2023	11/01/2023
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	11/01/2023	11/01/2023
					CT321	Long-term surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	11/01/2023	11/01/2023
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	11/01/2023	11/01/2023
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	11/01/2023	11/01/2023
					CT449	Generators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	11/01/2023	11/01/2023

							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	11/01/2023	11/01/2023
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	11/01/2023	11/01/2023
					CT981	Single use	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	11/01/2023	11/01/2023
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	11/01/2023	11/01/2023
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	11/01/2023	11/01/2023
					CT983	Surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	11/01/2023	11/01/2023
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	11/01/2023	11/01/2023
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	11/01/2023	11/01/2023
12349	2023		47265	Dual-chamber implantable pacemaker, rate-responsive	CT141	Electrodes/Leads and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	09/02/2023	09/02/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	09/02/2023	09/02/2023
					CT145	Cardiovascular devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	09/02/2023	09/02/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	09/02/2023	09/02/2023
					CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	09/02/2023	09/02/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	09/02/2023	09/02/2023
					CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	09/02/2023	09/02/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	09/02/2023	09/02/2023
					CT233	Surgical	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	09/02/2023	09/02/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	09/02/2023	09/02/2023
					CT2406	Implantable	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	09/02/2023	09/02/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	09/02/2023	09/02/2023
					CT2407	Active implantable	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	09/02/2023	09/02/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	09/02/2023	09/02/2023
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	09/02/2023	09/02/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	09/02/2023	09/02/2023
					CT2727	Active implant devices/ancillaries	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	09/02/2023	09/02/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	09/02/2023	09/02/2023
					CT2750	Implantable cardiac electrical stimulators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	09/02/2023	09/02/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	09/02/2023	09/02/2023
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	09/02/2023	09/02/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	09/02/2023	09/02/2023
					CT2985	Active	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	09/02/2023	09/02/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	09/02/2023	09/02/2023
					CT3001	Implantable metal	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	09/02/2023	09/02/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	09/02/2023	09/02/2023
					CT3002	Implantable synthetic polymer	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	09/02/2023	09/02/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	09/02/2023	09/02/2023
					CT3006	Implantable inorganic material	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	09/02/2023	09/02/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	09/02/2023	09/02/2023
					CT314	Electrically powered	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	09/02/2023	09/02/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	09/02/2023	09/02/2023
					CT321	Long-term surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	09/02/2023	09/02/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	09/02/2023	09/02/2023
					CT449	Generators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	09/02/2023	09/02/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	09/02/2023	09/02/2023
					CT981	Single use	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	09/02/2023	09/02/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	09/02/2023	09/02/2023
					CT983	Surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	09/02/2023	09/02/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	09/02/2023	09/02/2023
12350	2023		47265	Dual-chamber implantable pacemaker, rate-responsive	CT141	Electrodes/Leads and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	08/01/2023	08/01/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	08/01/2023	08/01/2023
					CT145	Cardiovascular devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	08/01/2023	08/01/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	08/01/2023	08/01/2023
					CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	08/01/2023	08/01/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	08/01/2023	08/01/2023
					CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	08/01/2023	08/01/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	08/01/2023	08/01/2023
					CT233	Surgical	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	08/01/2023	08/01/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	08/01/2023	08/01/2023
					CT2406	Implantable	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	08/01/2023	08/01/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	08/01/2023	08/01/2023
					CT2407	Active implantable	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	08/01/2023	08/01/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	08/01/2023	08/01/2023
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	08/01/2023	08/01/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	08/01/2023	08/01/2023
					CT2727	Active implant devices/ancillaries	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	08/01/2023	08/01/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	08/01/2023	08/01/2023
					CT2750	Implantable cardiac electrical stimulators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	08/01/2023	08/01/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	08/01/2023	08/01/2023
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	08/01/2023	08/01/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	08/01/2023	08/01/2023
					CT2985	Active	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	08/01/2023	08/01/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	08/01/2023	08/01/2023
					CT3001	Implantable metal	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	08/01/2023	08/01/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	08/01/2023	08/01/2023
					CT3002	Implantable synthetic polymer	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	08/01/2023	08/01/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	08/01/2023	08/01/2023
					CT3006	Implantable inorganic material	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	08/01/2023	08/01/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	08/01/2023	08/01/2023
					CT314	Electrically powered	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	08/01/2023	08/01/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	08/01/2023	08/01/2023
					CT321	Long-term surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	08/01/2023	08/01/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	08/01/2023	08/01/2023
					CT449	Generators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	08/01/2023	08/01/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	08/01/2023	08/01/2023
					CT981	Single use	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	08/01/2023	08/01/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	08/01/2023	08/01/2023
					CT983	Surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	08/01/2023	08/01/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	08/01/2023	08/01/2023

12351	2023		47265	Dual-chamber implantable pacemaker, rate-responsive	CT141	Electrodes/Leads and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	04/02/2023	04/02/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	04/02/2023	04/02/2023
					CT145	Cardiovascular devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	04/02/2023	04/02/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	04/02/2023	04/02/2023
					CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	04/02/2023	04/02/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	04/02/2023	04/02/2023
					CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	04/02/2023	04/02/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	04/02/2023	04/02/2023
					CT233	Surgical	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	04/02/2023	04/02/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	04/02/2023	04/02/2023
					CT2406	Implantable	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	04/02/2023	04/02/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	04/02/2023	04/02/2023
					CT2407	Active implantable	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	04/02/2023	04/02/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	04/02/2023	04/02/2023
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	04/02/2023	04/02/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	04/02/2023	04/02/2023
					CT2727	Active implant devices/ancillaries	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	04/02/2023	04/02/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	04/02/2023	04/02/2023
					CT2750	Implantable cardiac electrical stimulators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	04/02/2023	04/02/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	04/02/2023	04/02/2023
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	04/02/2023	04/02/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	04/02/2023	04/02/2023
					CT2985	Active	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	04/02/2023	04/02/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	04/02/2023	04/02/2023
					CT3001	Implantable metal	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	04/02/2023	04/02/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	04/02/2023	04/02/2023
					CT3002	Implantable synthetic polymer	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	04/02/2023	04/02/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	04/02/2023	04/02/2023
					CT3006	Implantable inorganic material	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	04/02/2023	04/02/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	04/02/2023	04/02/2023
					CT314	Electrically powered	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	04/02/2023	04/02/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	04/02/2023	04/02/2023
					CT321	Long-term surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	04/02/2023	04/02/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	04/02/2023	04/02/2023
					CT449	Generators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	04/02/2023	04/02/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	04/02/2023	04/02/2023
					CT981	Single use	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	04/02/2023	04/02/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	04/02/2023	04/02/2023
					CT983	Surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	04/02/2023	04/02/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	04/02/2023	04/02/2023
12352	2023	All other reportable incidents	47267	Single-chamber implantable pacemaker, rate-responsive	CT141	Electrodes/Leads and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F27	Problem Identified During Non-Clinical Procedure	A070504	Premature Discharge of Battery		
					CT145	Cardiovascular devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F27	Problem Identified During Non-Clinical Procedure	A070504	Premature Discharge of Battery		
					CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F27	Problem Identified During Non-Clinical Procedure	A070504	Premature Discharge of Battery		
					CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F27	Problem Identified During Non-Clinical Procedure	A070504	Premature Discharge of Battery		
					CT233	Surgical	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F27	Problem Identified During Non-Clinical Procedure	A070504	Premature Discharge of Battery		
					CT2406	Implantable	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F27	Problem Identified During Non-Clinical Procedure	A070504	Premature Discharge of Battery		
					CT2407	Active implantable	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F27	Problem Identified During Non-Clinical Procedure	A070504	Premature Discharge of Battery		
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F27	Problem Identified During Non-Clinical Procedure	A070504	Premature Discharge of Battery		
					CT2727	Active implant devices/ancillaries	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F27	Problem Identified During Non-Clinical Procedure	A070504	Premature Discharge of Battery		
					CT2750	Implantable cardiac electrical stimulators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F27	Problem Identified During Non-Clinical Procedure	A070504	Premature Discharge of Battery		
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F27	Problem Identified During Non-Clinical Procedure	A070504	Premature Discharge of Battery		
					CT2985	Active	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F27	Problem Identified During Non-Clinical Procedure	A070504	Premature Discharge of Battery		
					CT3001	Implantable metal	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F27	Problem Identified During Non-Clinical Procedure	A070504	Premature Discharge of Battery		
					CT3002	Implantable synthetic polymer	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F27	Problem Identified During Non-Clinical Procedure	A070504	Premature Discharge of Battery		
					CT3006	Implantable inorganic material	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F27	Problem Identified During Non-Clinical Procedure	A070504	Premature Discharge of Battery		
					CT314	Electrically powered	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F27	Problem Identified During Non-Clinical Procedure	A070504	Premature Discharge of Battery		
					CT321	Long-term surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F27	Problem Identified During Non-Clinical Procedure	A070504	Premature Discharge of Battery		
					CT449	Generators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F27	Problem Identified During Non-Clinical Procedure	A070504	Premature Discharge of Battery		
					CT981	Single use	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F27	Problem Identified During Non-Clinical Procedure	A070504	Premature Discharge of Battery		
					CT983	Surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F27	Problem Identified During Non-Clinical Procedure	A070504	Premature Discharge of Battery		
12353	2023	All other reportable incidents	47263	Cardiac resynchronization therapy implantable pacemaker	CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A0908	Incorrect, Inadequate or Imprecise Result or Readings		
							F11	Non-Serious Injury/ Illness/ Impairment	A0908	Incorrect, Inadequate or Imprecise Result or Readings		
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A0908	Incorrect, Inadequate or Imprecise Result or Readings		
							F11	Non-Serious Injury/ Illness/ Impairment	A0908	Incorrect, Inadequate or Imprecise Result or Readings		

					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A0908	Incorrect, Inadequate or Imprecise Result or Readings		
							F11	Non-Serious Injury/ Illness/ Impairment	A0908	Incorrect, Inadequate or Imprecise Result or Readings		
12354	2023	All other reportable incidents	47267	Single-chamber implantable pacemaker, rate-responsive	CT141	Electrodes/Leads and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem	29/11/2022	29/11/2022
									A070504	Premature Discharge of Battery	29/11/2022	29/11/2022
							F08	Hospitalization or Prolonged Hospitalization	A0705	Battery Problem	29/11/2022	29/11/2022
									A070504	Premature Discharge of Battery	29/11/2022	29/11/2022
							F1905	Device Revision or Replacement	A0705	Battery Problem	29/11/2022	29/11/2022
									A070504	Premature Discharge of Battery	29/11/2022	29/11/2022
					CT145	Cardiovascular devices	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem	29/11/2022	29/11/2022
									A070504	Premature Discharge of Battery	29/11/2022	29/11/2022
							F08	Hospitalization or Prolonged Hospitalization	A0705	Battery Problem	29/11/2022	29/11/2022
									A070504	Premature Discharge of Battery	29/11/2022	29/11/2022
							F1905	Device Revision or Replacement	A0705	Battery Problem	29/11/2022	29/11/2022
									A070504	Premature Discharge of Battery	29/11/2022	29/11/2022
					CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem	29/11/2022	29/11/2022
									A070504	Premature Discharge of Battery	29/11/2022	29/11/2022
							F08	Hospitalization or Prolonged Hospitalization	A0705	Battery Problem	29/11/2022	29/11/2022
									A070504	Premature Discharge of Battery	29/11/2022	29/11/2022
							F1905	Device Revision or Replacement	A0705	Battery Problem	29/11/2022	29/11/2022
									A070504	Premature Discharge of Battery	29/11/2022	29/11/2022
					CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem	29/11/2022	29/11/2022
									A070504	Premature Discharge of Battery	29/11/2022	29/11/2022
							F08	Hospitalization or Prolonged Hospitalization	A0705	Battery Problem	29/11/2022	29/11/2022
									A070504	Premature Discharge of Battery	29/11/2022	29/11/2022
							F1905	Device Revision or Replacement	A0705	Battery Problem	29/11/2022	29/11/2022
									A070504	Premature Discharge of Battery	29/11/2022	29/11/2022
					CT233	Surgical	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem	29/11/2022	29/11/2022
									A070504	Premature Discharge of Battery	29/11/2022	29/11/2022
							F08	Hospitalization or Prolonged Hospitalization	A0705	Battery Problem	29/11/2022	29/11/2022
									A070504	Premature Discharge of Battery	29/11/2022	29/11/2022
							F1905	Device Revision or Replacement	A0705	Battery Problem	29/11/2022	29/11/2022
									A070504	Premature Discharge of Battery	29/11/2022	29/11/2022
					CT2406	Implantable	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem	29/11/2022	29/11/2022
									A070504	Premature Discharge of Battery	29/11/2022	29/11/2022
							F08	Hospitalization or Prolonged Hospitalization	A0705	Battery Problem	29/11/2022	29/11/2022
									A070504	Premature Discharge of Battery	29/11/2022	29/11/2022
							F1905	Device Revision or Replacement	A0705	Battery Problem	29/11/2022	29/11/2022
									A070504	Premature Discharge of Battery	29/11/2022	29/11/2022
					CT2407	Active implantable	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem	29/11/2022	29/11/2022
									A070504	Premature Discharge of Battery	29/11/2022	29/11/2022
							F08	Hospitalization or Prolonged Hospitalization	A0705	Battery Problem	29/11/2022	29/11/2022
									A070504	Premature Discharge of Battery	29/11/2022	29/11/2022
							F1905	Device Revision or Replacement	A0705	Battery Problem	29/11/2022	29/11/2022
									A070504	Premature Discharge of Battery	29/11/2022	29/11/2022
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem	29/11/2022	29/11/2022
									A070504	Premature Discharge of Battery	29/11/2022	29/11/2022
							F08	Hospitalization or Prolonged Hospitalization	A0705	Battery Problem	29/11/2022	29/11/2022
									A070504	Premature Discharge of Battery	29/11/2022	29/11/2022
							F1905	Device Revision or Replacement	A0705	Battery Problem	29/11/2022	29/11/2022
									A070504	Premature Discharge of Battery	29/11/2022	29/11/2022
					CT2727	Active implant devices/ancillaries	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem	29/11/2022	29/11/2022
									A070504	Premature Discharge of Battery	29/11/2022	29/11/2022
							F08	Hospitalization or Prolonged Hospitalization	A0705	Battery Problem	29/11/2022	29/11/2022
									A070504	Premature Discharge of Battery	29/11/2022	29/11/2022
							F1905	Device Revision or Replacement	A0705	Battery Problem	29/11/2022	29/11/2022
									A070504	Premature Discharge of Battery	29/11/2022	29/11/2022
					CT2750	Implantable cardiac electrical stimulators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem	29/11/2022	29/11/2022
									A070504	Premature Discharge of Battery	29/11/2022	29/11/2022
							F08	Hospitalization or Prolonged Hospitalization	A0705	Battery Problem	29/11/2022	29/11/2022
									A070504	Premature Discharge of Battery	29/11/2022	29/11/2022
							F1905	Device Revision or Replacement	A0705	Battery Problem	29/11/2022	29/11/2022
									A070504	Premature Discharge of Battery	29/11/2022	29/11/2022
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem	29/11/2022	29/11/2022
									A070504	Premature Discharge of Battery	29/11/2022	29/11/2022
							F08	Hospitalization or Prolonged Hospitalization	A0705	Battery Problem	29/11/2022	29/11/2022
									A070504	Premature Discharge of Battery	29/11/2022	29/11/2022
							F1905	Device Revision or Replacement	A0705	Battery Problem	29/11/2022	29/11/2022
									A070504	Premature Discharge of Battery	29/11/2022	29/11/2022
					CT2985	Active	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem	29/11/2022	29/11/2022
									A070504	Premature Discharge of Battery	29/11/2022	29/11/2022
							F08	Hospitalization or Prolonged Hospitalization	A0705	Battery Problem	29/11/2022	29/11/2022
									A070504	Premature Discharge of Battery	29/11/2022	29/11/2022
							F1905	Device Revision or Replacement	A0705	Battery Problem	29/11/2022	29/11/2022
									A070504	Premature Discharge of Battery	29/11/2022	29/11/2022
					CT3001	Implantable metal	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem	29/11/2022	29/11/2022
									A070504	Premature Discharge of Battery	29/11/2022	29/11/2022
							F08	Hospitalization or Prolonged Hospitalization	A0705	Battery Problem	29/11/2022	29/11/2022
									A070504	Premature Discharge of Battery	29/11/2022	29/11/2022
							F1905	Device Revision or Replacement	A0705	Battery Problem	29/11/2022	29/11/2022
									A070504	Premature Discharge of Battery	29/11/2022	29/11/2022
					CT3002	Implantable synthetic polymer	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem	29/11/2022	29/11/2022
									A070504	Premature Discharge of Battery	29/11/2022	29/11/2022
							F08	Hospitalization or Prolonged Hospitalization	A0705	Battery Problem	29/11/2022	29/11/2022
									A070504	Premature Discharge of Battery	29/11/2022	29/11/2022
							F1905	Device Revision or Replacement	A0705	Battery Problem	29/11/2022	29/11/2022
									A070504	Premature Discharge of Battery	29/11/2022	29/11/2022

					CT3006	Implantable inorganic material	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem	29/11/2022	29/11/2022
									A070504	Premature Discharge of Battery	29/11/2022	29/11/2022
							F08	Hospitalization or Prolonged Hospitalization	A0705	Battery Problem	29/11/2022	29/11/2022
									A070504	Premature Discharge of Battery	29/11/2022	29/11/2022
							F1905	Device Revision or Replacement	A0705	Battery Problem	29/11/2022	29/11/2022
									A070504	Premature Discharge of Battery	29/11/2022	29/11/2022
					CT314	Electrically powered	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem	29/11/2022	29/11/2022
									A070504	Premature Discharge of Battery	29/11/2022	29/11/2022
							F08	Hospitalization or Prolonged Hospitalization	A0705	Battery Problem	29/11/2022	29/11/2022
									A070504	Premature Discharge of Battery	29/11/2022	29/11/2022
							F1905	Device Revision or Replacement	A0705	Battery Problem	29/11/2022	29/11/2022
									A070504	Premature Discharge of Battery	29/11/2022	29/11/2022
					CT321	Long-term surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem	29/11/2022	29/11/2022
									A070504	Premature Discharge of Battery	29/11/2022	29/11/2022
							F08	Hospitalization or Prolonged Hospitalization	A0705	Battery Problem	29/11/2022	29/11/2022
									A070504	Premature Discharge of Battery	29/11/2022	29/11/2022
							F1905	Device Revision or Replacement	A0705	Battery Problem	29/11/2022	29/11/2022
									A070504	Premature Discharge of Battery	29/11/2022	29/11/2022
					CT449	Generators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem	29/11/2022	29/11/2022
									A070504	Premature Discharge of Battery	29/11/2022	29/11/2022
							F08	Hospitalization or Prolonged Hospitalization	A0705	Battery Problem	29/11/2022	29/11/2022
									A070504	Premature Discharge of Battery	29/11/2022	29/11/2022
							F1905	Device Revision or Replacement	A0705	Battery Problem	29/11/2022	29/11/2022
									A070504	Premature Discharge of Battery	29/11/2022	29/11/2022
					CT981	Single use	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem	29/11/2022	29/11/2022
									A070504	Premature Discharge of Battery	29/11/2022	29/11/2022
							F08	Hospitalization or Prolonged Hospitalization	A0705	Battery Problem	29/11/2022	29/11/2022
									A070504	Premature Discharge of Battery	29/11/2022	29/11/2022
							F1905	Device Revision or Replacement	A0705	Battery Problem	29/11/2022	29/11/2022
									A070504	Premature Discharge of Battery	29/11/2022	29/11/2022
					CT983	Surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem	29/11/2022	29/11/2022
									A070504	Premature Discharge of Battery	29/11/2022	29/11/2022
							F08	Hospitalization or Prolonged Hospitalization	A0705	Battery Problem	29/11/2022	29/11/2022
									A070504	Premature Discharge of Battery	29/11/2022	29/11/2022
							F1905	Device Revision or Replacement	A0705	Battery Problem	29/11/2022	29/11/2022
									A070504	Premature Discharge of Battery	29/11/2022	29/11/2022
12355	2023		47265	Dual-chamber implantable pacemaker, rate-responsive	CT141	Electrodes/Leads and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem		
							F1901	Additional Surgery	A24	No Apparent Device Problem		
					CT145	Cardiovascular devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem		
							F1901	Additional Surgery	A24	No Apparent Device Problem		
					CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem		
							F1901	Additional Surgery	A24	No Apparent Device Problem		
					CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem		
							F1901	Additional Surgery	A24	No Apparent Device Problem		
					CT233	Surgical	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem		
							F1901	Additional Surgery	A24	No Apparent Device Problem		
					CT2406	Implantable	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem		
							F1901	Additional Surgery	A24	No Apparent Device Problem		
					CT2407	Active implantable	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem		
							F1901	Additional Surgery	A24	No Apparent Device Problem		
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem		
							F1901	Additional Surgery	A24	No Apparent Device Problem		
					CT2727	Active implant devices/ancillaries	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem		
							F1901	Additional Surgery	A24	No Apparent Device Problem		
					CT2750	Implantable cardiac electrical stimulators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem		
							F1901	Additional Surgery	A24	No Apparent Device Problem		
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem		
							F1901	Additional Surgery	A24	No Apparent Device Problem		
					CT2985	Active	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem		
							F1901	Additional Surgery	A24	No Apparent Device Problem		
					CT3001	Implantable metal	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem		
							F1901	Additional Surgery	A24	No Apparent Device Problem		
					CT3002	Implantable synthetic polymer	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem		
							F1901	Additional Surgery	A24	No Apparent Device Problem		
					CT3006	Implantable inorganic material	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem		
							F1901	Additional Surgery	A24	No Apparent Device Problem		
					CT314	Electrically powered	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem		
							F1901	Additional Surgery	A24	No Apparent Device Problem		
					CT321	Long-term surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem		
							F1901	Additional Surgery	A24	No Apparent Device Problem		
					CT449	Generators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem		
							F1901	Additional Surgery	A24	No Apparent Device Problem		
					CT981	Single use	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem		
							F1901	Additional Surgery	A24	No Apparent Device Problem		
					CT983	Surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem		
							F1901	Additional Surgery	A24	No Apparent Device Problem		
12356	2023		47265	Dual-chamber implantable pacemaker, rate-responsive	CT141	Electrodes/Leads and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	31/01/2023	31/01/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	31/01/2023	31/01/2023
					CT145	Cardiovascular devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	31/01/2023	31/01/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	31/01/2023	31/01/2023
					CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	31/01/2023	31/01/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	31/01/2023	31/01/2023
					CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	31/01/2023	31/01/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	31/01/2023	31/01/2023
					CT233	Surgical	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	31/01/2023	31/01/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	31/01/2023	31/01/2023
					CT2406	Implantable	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	31/01/2023	31/01/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	31/01/2023	31/01/2023

					CT2407	Active implantable	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	31/01/2023	31/01/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	31/01/2023	31/01/2023
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	31/01/2023	31/01/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	31/01/2023	31/01/2023
					CT2727	Active implant devices/ancillaries	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	31/01/2023	31/01/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	31/01/2023	31/01/2023
					CT2750	Implantable cardiac electrical stimulators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	31/01/2023	31/01/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	31/01/2023	31/01/2023
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	31/01/2023	31/01/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	31/01/2023	31/01/2023
					CT2985	Active	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	31/01/2023	31/01/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	31/01/2023	31/01/2023
					CT3001	Implantable metal	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	31/01/2023	31/01/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	31/01/2023	31/01/2023
					CT3002	Implantable synthetic polymer	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	31/01/2023	31/01/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	31/01/2023	31/01/2023
					CT3006	Implantable inorganic material	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	31/01/2023	31/01/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	31/01/2023	31/01/2023
					CT314	Electrically powered	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	31/01/2023	31/01/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	31/01/2023	31/01/2023
					CT321	Long-term surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	31/01/2023	31/01/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	31/01/2023	31/01/2023
					CT449	Generators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	31/01/2023	31/01/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	31/01/2023	31/01/2023
					CT981	Single use	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	31/01/2023	31/01/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	31/01/2023	31/01/2023
					CT983	Surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	31/01/2023	31/01/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	31/01/2023	31/01/2023
12357	2023	All other reportable incidents	47265	Dual-chamber implantable pacemaker, rate-responsive	CT141	Electrodes/Leads and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	28/01/2023	28/01/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	28/01/2023	28/01/2023
					CT145	Cardiovascular devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	28/01/2023	28/01/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	28/01/2023	28/01/2023
					CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	28/01/2023	28/01/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	28/01/2023	28/01/2023
					CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	28/01/2023	28/01/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	28/01/2023	28/01/2023
					CT233	Surgical	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	28/01/2023	28/01/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	28/01/2023	28/01/2023
					CT2406	Implantable	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	28/01/2023	28/01/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	28/01/2023	28/01/2023
					CT2407	Active implantable	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	28/01/2023	28/01/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	28/01/2023	28/01/2023
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	28/01/2023	28/01/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	28/01/2023	28/01/2023
					CT2727	Active implant devices/ancillaries	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	28/01/2023	28/01/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	28/01/2023	28/01/2023
					CT2750	Implantable cardiac electrical stimulators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	28/01/2023	28/01/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	28/01/2023	28/01/2023
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	28/01/2023	28/01/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	28/01/2023	28/01/2023
					CT2985	Active	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	28/01/2023	28/01/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	28/01/2023	28/01/2023
					CT3001	Implantable metal	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	28/01/2023	28/01/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	28/01/2023	28/01/2023
					CT3002	Implantable synthetic polymer	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	28/01/2023	28/01/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	28/01/2023	28/01/2023
					CT3006	Implantable inorganic material	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	28/01/2023	28/01/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	28/01/2023	28/01/2023
					CT314	Electrically powered	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	28/01/2023	28/01/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	28/01/2023	28/01/2023
					CT321	Long-term surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	28/01/2023	28/01/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	28/01/2023	28/01/2023
					CT449	Generators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	28/01/2023	28/01/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	28/01/2023	28/01/2023
					CT981	Single use	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	28/01/2023	28/01/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	28/01/2023	28/01/2023
					CT983	Surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	28/01/2023	28/01/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	28/01/2023	28/01/2023
12358	2023		47265	Dual-chamber implantable pacemaker, rate-responsive	CT141	Electrodes/Leads and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	27/01/2023	27/01/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	27/01/2023	27/01/2023
					CT145	Cardiovascular devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	27/01/2023	27/01/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	27/01/2023	27/01/2023
					CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	27/01/2023	27/01/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	27/01/2023	27/01/2023
					CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	27/01/2023	27/01/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	27/01/2023	27/01/2023
					CT233	Surgical	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	27/01/2023	27/01/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	27/01/2023	27/01/2023
					CT2406	Implantable	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	27/01/2023	27/01/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	27/01/2023	27/01/2023
					CT2407	Active implantable	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	27/01/2023	27/01/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	27/01/2023	27/01/2023
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	27/01/2023	27/01/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	27/01/2023	27/01/2023
					CT2727	Active implant devices/ancillaries	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	27/01/2023	27/01/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	27/01/2023	27/01/2023
					CT2750	Implantable cardiac electrical stimulators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	27/01/2023	27/01/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	27/01/2023	27/01/2023

					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	27/01/2023	27/01/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	27/01/2023	27/01/2023
					CT2985	Active	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	27/01/2023	27/01/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	27/01/2023	27/01/2023
					CT3001	Implantable metal	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	27/01/2023	27/01/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	27/01/2023	27/01/2023
					CT3002	Implantable synthetic polymer	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	27/01/2023	27/01/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	27/01/2023	27/01/2023
					CT3006	Implantable inorganic material	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	27/01/2023	27/01/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	27/01/2023	27/01/2023
					CT314	Electrically powered	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	27/01/2023	27/01/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	27/01/2023	27/01/2023
					CT321	Long-term surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	27/01/2023	27/01/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	27/01/2023	27/01/2023
					CT449	Generators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	27/01/2023	27/01/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	27/01/2023	27/01/2023
					CT981	Single use	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	27/01/2023	27/01/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	27/01/2023	27/01/2023
					CT983	Surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	27/01/2023	27/01/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	27/01/2023	27/01/2023
12359	2023		47265	Dual-chamber implantable pacemaker, rate-responsive	CT141	Electrodes/Leads and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	11/02/2023	11/02/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	11/02/2023	11/02/2023
					CT145	Cardiovascular devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	11/02/2023	11/02/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	11/02/2023	11/02/2023
					CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	11/02/2023	11/02/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	11/02/2023	11/02/2023
					CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	11/02/2023	11/02/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	11/02/2023	11/02/2023
					CT233	Surgical	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	11/02/2023	11/02/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	11/02/2023	11/02/2023
					CT2406	Implantable	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	11/02/2023	11/02/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	11/02/2023	11/02/2023
					CT2407	Active implantable	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	11/02/2023	11/02/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	11/02/2023	11/02/2023
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	11/02/2023	11/02/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	11/02/2023	11/02/2023
					CT2727	Active implant devices/ancillaries	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	11/02/2023	11/02/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	11/02/2023	11/02/2023
					CT2750	Implantable cardiac electrical stimulators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	11/02/2023	11/02/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	11/02/2023	11/02/2023
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	11/02/2023	11/02/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	11/02/2023	11/02/2023
					CT2985	Active	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	11/02/2023	11/02/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	11/02/2023	11/02/2023
					CT3001	Implantable metal	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	11/02/2023	11/02/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	11/02/2023	11/02/2023
					CT3002	Implantable synthetic polymer	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	11/02/2023	11/02/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	11/02/2023	11/02/2023
					CT3006	Implantable inorganic material	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	11/02/2023	11/02/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	11/02/2023	11/02/2023
					CT314	Electrically powered	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	11/02/2023	11/02/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	11/02/2023	11/02/2023
					CT321	Long-term surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	11/02/2023	11/02/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	11/02/2023	11/02/2023
					CT449	Generators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	11/02/2023	11/02/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	11/02/2023	11/02/2023
					CT981	Single use	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	11/02/2023	11/02/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	11/02/2023	11/02/2023
					CT983	Surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	11/02/2023	11/02/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	11/02/2023	11/02/2023
12360	2023	All other reportable incidents	47265	Dual-chamber implantable pacemaker, rate-responsive	CT141	Electrodes/Leads and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	29/01/2023	29/01/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	29/01/2023	29/01/2023
					CT145	Cardiovascular devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	29/01/2023	29/01/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	29/01/2023	29/01/2023
					CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	29/01/2023	29/01/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	29/01/2023	29/01/2023
					CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	29/01/2023	29/01/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	29/01/2023	29/01/2023
					CT233	Surgical	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	29/01/2023	29/01/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	29/01/2023	29/01/2023
					CT2406	Implantable	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	29/01/2023	29/01/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	29/01/2023	29/01/2023
					CT2407	Active implantable	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	29/01/2023	29/01/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	29/01/2023	29/01/2023
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	29/01/2023	29/01/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	29/01/2023	29/01/2023
					CT2727	Active implant devices/ancillaries	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	29/01/2023	29/01/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	29/01/2023	29/01/2023
					CT2750	Implantable cardiac electrical stimulators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	29/01/2023	29/01/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	29/01/2023	29/01/2023
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	29/01/2023	29/01/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	29/01/2023	29/01/2023
					CT2985	Active	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	29/01/2023	29/01/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	29/01/2023	29/01/2023
					CT3001	Implantable metal	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	29/01/2023	29/01/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	29/01/2023	29/01/2023
					CT3002	Implantable synthetic polymer	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	29/01/2023	29/01/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	29/01/2023	29/01/2023

					CT3006	Implantable inorganic material	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	29/01/2023	29/01/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	29/01/2023	29/01/2023
					CT314	Electrically powered	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	29/01/2023	29/01/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	29/01/2023	29/01/2023
					CT321	Long-term surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	29/01/2023	29/01/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	29/01/2023	29/01/2023
					CT449	Generators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	29/01/2023	29/01/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	29/01/2023	29/01/2023
					CT981	Single use	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	29/01/2023	29/01/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	29/01/2023	29/01/2023
					CT983	Surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	29/01/2023	29/01/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	29/01/2023	29/01/2023
12361	2023	All other reportable incidents	47265	Dual-chamber implantable pacemaker, rate-responsive	CT141	Electrodes/Leads and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	31/01/2023	31/01/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	31/01/2023	31/01/2023
					CT145	Cardiovascular devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	31/01/2023	31/01/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	31/01/2023	31/01/2023
					CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	31/01/2023	31/01/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	31/01/2023	31/01/2023
					CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	31/01/2023	31/01/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	31/01/2023	31/01/2023
					CT233	Surgical	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	31/01/2023	31/01/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	31/01/2023	31/01/2023
					CT2406	Implantable	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	31/01/2023	31/01/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	31/01/2023	31/01/2023
					CT2407	Active implantable	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	31/01/2023	31/01/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	31/01/2023	31/01/2023
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	31/01/2023	31/01/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	31/01/2023	31/01/2023
					CT2727	Active implant devices/ancillaries	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	31/01/2023	31/01/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	31/01/2023	31/01/2023
					CT2750	Implantable cardiac electrical stimulators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	31/01/2023	31/01/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	31/01/2023	31/01/2023
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	31/01/2023	31/01/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	31/01/2023	31/01/2023
					CT2985	Active	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	31/01/2023	31/01/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	31/01/2023	31/01/2023
					CT3001	Implantable metal	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	31/01/2023	31/01/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	31/01/2023	31/01/2023
					CT3002	Implantable synthetic polymer	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	31/01/2023	31/01/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	31/01/2023	31/01/2023
					CT3006	Implantable inorganic material	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	31/01/2023	31/01/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	31/01/2023	31/01/2023
					CT314	Electrically powered	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	31/01/2023	31/01/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	31/01/2023	31/01/2023
					CT321	Long-term surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	31/01/2023	31/01/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	31/01/2023	31/01/2023
					CT449	Generators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	31/01/2023	31/01/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	31/01/2023	31/01/2023
					CT981	Single use	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	31/01/2023	31/01/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	31/01/2023	31/01/2023
					CT983	Surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	31/01/2023	31/01/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	31/01/2023	31/01/2023
12362	2023		47265	Dual-chamber implantable pacemaker, rate-responsive	CT141	Electrodes/Leads and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	19/01/2023	19/01/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	19/01/2023	19/01/2023
					CT145	Cardiovascular devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	19/01/2023	19/01/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	19/01/2023	19/01/2023
					CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	19/01/2023	19/01/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	19/01/2023	19/01/2023
					CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	19/01/2023	19/01/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	19/01/2023	19/01/2023
					CT233	Surgical	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	19/01/2023	19/01/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	19/01/2023	19/01/2023
					CT2406	Implantable	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	19/01/2023	19/01/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	19/01/2023	19/01/2023
					CT2407	Active implantable	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	19/01/2023	19/01/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	19/01/2023	19/01/2023
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	19/01/2023	19/01/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	19/01/2023	19/01/2023
					CT2727	Active implant devices/ancillaries	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	19/01/2023	19/01/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	19/01/2023	19/01/2023
					CT2750	Implantable cardiac electrical stimulators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	19/01/2023	19/01/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	19/01/2023	19/01/2023
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	19/01/2023	19/01/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	19/01/2023	19/01/2023
					CT2985	Active	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	19/01/2023	19/01/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	19/01/2023	19/01/2023
					CT3001	Implantable metal	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	19/01/2023	19/01/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	19/01/2023	19/01/2023
					CT3002	Implantable synthetic polymer	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	19/01/2023	19/01/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	19/01/2023	19/01/2023
					CT3006	Implantable inorganic material	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	19/01/2023	19/01/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	19/01/2023	19/01/2023
					CT314	Electrically powered	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	19/01/2023	19/01/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	19/01/2023	19/01/2023
					CT321	Long-term surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	19/01/2023	19/01/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	19/01/2023	19/01/2023
					CT449	Generators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	19/01/2023	19/01/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	19/01/2023	19/01/2023

					CT981	Single use	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	19/01/2023	19/01/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	19/01/2023	19/01/2023
					CT983	Surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	19/01/2023	19/01/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	19/01/2023	19/01/2023
12363	2023		47265	Dual-chamber implantable pacemaker, rate-responsive	CT141	Electrodes/Leads and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	11/02/2023	11/02/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	11/02/2023	11/02/2023
					CT145	Cardiovascular devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	11/02/2023	11/02/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	11/02/2023	11/02/2023
					CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	11/02/2023	11/02/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	11/02/2023	11/02/2023
					CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	11/02/2023	11/02/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	11/02/2023	11/02/2023
					CT233	Surgical	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	11/02/2023	11/02/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	11/02/2023	11/02/2023
					CT2406	Implantable	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	11/02/2023	11/02/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	11/02/2023	11/02/2023
					CT2407	Active implantable	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	11/02/2023	11/02/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	11/02/2023	11/02/2023
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	11/02/2023	11/02/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	11/02/2023	11/02/2023
					CT2727	Active implant devices/ancillaries	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	11/02/2023	11/02/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	11/02/2023	11/02/2023
					CT2750	Implantable cardiac electrical stimulators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	11/02/2023	11/02/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	11/02/2023	11/02/2023
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	11/02/2023	11/02/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	11/02/2023	11/02/2023
					CT2985	Active	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	11/02/2023	11/02/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	11/02/2023	11/02/2023
					CT3001	Implantable metal	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	11/02/2023	11/02/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	11/02/2023	11/02/2023
					CT3002	Implantable synthetic polymer	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	11/02/2023	11/02/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	11/02/2023	11/02/2023
					CT3006	Implantable inorganic material	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	11/02/2023	11/02/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	11/02/2023	11/02/2023
					CT314	Electrically powered	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	11/02/2023	11/02/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	11/02/2023	11/02/2023
					CT321	Long-term surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	11/02/2023	11/02/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	11/02/2023	11/02/2023
					CT449	Generators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	11/02/2023	11/02/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	11/02/2023	11/02/2023
					CT981	Single use	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	11/02/2023	11/02/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	11/02/2023	11/02/2023
					CT983	Surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	11/02/2023	11/02/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	11/02/2023	11/02/2023
12364	2023		47265	Dual-chamber implantable pacemaker, rate-responsive	CT141	Electrodes/Leads and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	11/02/2023	11/02/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	11/02/2023	11/02/2023
					CT145	Cardiovascular devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	11/02/2023	11/02/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	11/02/2023	11/02/2023
					CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	11/02/2023	11/02/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	11/02/2023	11/02/2023
					CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	11/02/2023	11/02/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	11/02/2023	11/02/2023
					CT233	Surgical	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	11/02/2023	11/02/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	11/02/2023	11/02/2023
					CT2406	Implantable	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	11/02/2023	11/02/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	11/02/2023	11/02/2023
					CT2407	Active implantable	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	11/02/2023	11/02/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	11/02/2023	11/02/2023
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	11/02/2023	11/02/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	11/02/2023	11/02/2023
					CT2727	Active implant devices/ancillaries	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	11/02/2023	11/02/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	11/02/2023	11/02/2023
					CT2750	Implantable cardiac electrical stimulators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	11/02/2023	11/02/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	11/02/2023	11/02/2023
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	11/02/2023	11/02/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	11/02/2023	11/02/2023
					CT2985	Active	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	11/02/2023	11/02/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	11/02/2023	11/02/2023
					CT3001	Implantable metal	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	11/02/2023	11/02/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	11/02/2023	11/02/2023
					CT3002	Implantable synthetic polymer	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	11/02/2023	11/02/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	11/02/2023	11/02/2023
					CT3006	Implantable inorganic material	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	11/02/2023	11/02/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	11/02/2023	11/02/2023
					CT314	Electrically powered	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	11/02/2023	11/02/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	11/02/2023	11/02/2023
					CT321	Long-term surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	11/02/2023	11/02/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	11/02/2023	11/02/2023
					CT449	Generators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	11/02/2023	11/02/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	11/02/2023	11/02/2023
					CT981	Single use	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	11/02/2023	11/02/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	11/02/2023	11/02/2023
					CT983	Surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	11/02/2023	11/02/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	11/02/2023	11/02/2023
12365	2023		47265	Dual-chamber implantable pacemaker, rate-responsive	CT141	Electrodes/Leads and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	13/02/2023	13/02/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	13/02/2023	13/02/2023
					CT145	Cardiovascular devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	13/02/2023	13/02/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	13/02/2023	13/02/2023

					CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	13/02/2023	13/02/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	13/02/2023	13/02/2023
					CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	13/02/2023	13/02/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	13/02/2023	13/02/2023
					CT233	Surgical	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	13/02/2023	13/02/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	13/02/2023	13/02/2023
					CT2406	Implantable	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	13/02/2023	13/02/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	13/02/2023	13/02/2023
					CT2407	Active implantable	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	13/02/2023	13/02/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	13/02/2023	13/02/2023
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	13/02/2023	13/02/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	13/02/2023	13/02/2023
					CT2727	Active implant devices/ancillaries	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	13/02/2023	13/02/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	13/02/2023	13/02/2023
					CT2750	Implantable cardiac electrical stimulators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	13/02/2023	13/02/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	13/02/2023	13/02/2023
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	13/02/2023	13/02/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	13/02/2023	13/02/2023
					CT2985	Active	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	13/02/2023	13/02/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	13/02/2023	13/02/2023
					CT3001	Implantable metal	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	13/02/2023	13/02/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	13/02/2023	13/02/2023
					CT3002	Implantable synthetic polymer	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	13/02/2023	13/02/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	13/02/2023	13/02/2023
					CT3006	Implantable inorganic material	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	13/02/2023	13/02/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	13/02/2023	13/02/2023
					CT314	Electrically powered	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	13/02/2023	13/02/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	13/02/2023	13/02/2023
					CT321	Long-term surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	13/02/2023	13/02/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	13/02/2023	13/02/2023
					CT449	Generators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	13/02/2023	13/02/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	13/02/2023	13/02/2023
					CT981	Single use	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	13/02/2023	13/02/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	13/02/2023	13/02/2023
					CT983	Surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	13/02/2023	13/02/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	13/02/2023	13/02/2023
12366	2023	All other reportable incidents	37265	Dual-chamber implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E2104	Electric Shock	A0712	Pacing Problem		
									A071302	Inappropriate/Inadequate Shock/Stimulation		
							F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A0712	Pacing Problem		
									A071302	Inappropriate/Inadequate Shock/Stimulation		
12367	2023		47265	Dual-chamber implantable pacemaker, rate-responsive	CT141	Electrodes/Leads and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	15/02/2023	15/02/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	15/02/2023	15/02/2023
					CT145	Cardiovascular devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	15/02/2023	15/02/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	15/02/2023	15/02/2023
					CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	15/02/2023	15/02/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	15/02/2023	15/02/2023
					CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	15/02/2023	15/02/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	15/02/2023	15/02/2023
					CT233	Surgical	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	15/02/2023	15/02/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	15/02/2023	15/02/2023
					CT2406	Implantable	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	15/02/2023	15/02/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	15/02/2023	15/02/2023
					CT2407	Active implantable	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	15/02/2023	15/02/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	15/02/2023	15/02/2023
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	15/02/2023	15/02/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	15/02/2023	15/02/2023
					CT2727	Active implant devices/ancillaries	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	15/02/2023	15/02/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	15/02/2023	15/02/2023
					CT2750	Implantable cardiac electrical stimulators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	15/02/2023	15/02/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	15/02/2023	15/02/2023
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	15/02/2023	15/02/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	15/02/2023	15/02/2023
					CT2985	Active	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	15/02/2023	15/02/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	15/02/2023	15/02/2023
					CT3001	Implantable metal	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	15/02/2023	15/02/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	15/02/2023	15/02/2023
					CT3002	Implantable synthetic polymer	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	15/02/2023	15/02/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	15/02/2023	15/02/2023
					CT3006	Implantable inorganic material	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	15/02/2023	15/02/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	15/02/2023	15/02/2023
					CT314	Electrically powered	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	15/02/2023	15/02/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	15/02/2023	15/02/2023
					CT321	Long-term surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	15/02/2023	15/02/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	15/02/2023	15/02/2023
					CT449	Generators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	15/02/2023	15/02/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	15/02/2023	15/02/2023
					CT981	Single use	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	15/02/2023	15/02/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	15/02/2023	15/02/2023
					CT983	Surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	15/02/2023	15/02/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	15/02/2023	15/02/2023
12368	2023	All other reportable incidents	63188	Left-ventricular wireless pacing system sensor/transmitter	CT141	Electrodes/Leads and associated devices	F24	Insufficient Health Impact Information	A0708	Power Problem		
					CT145	Cardiovascular devices	F24	Insufficient Health Impact Information	A0708	Power Problem		
					CT1577	Wireless monitoring system transmitters	F24	Insufficient Health Impact Information	A0708	Power Problem		
					CT1684	Cardiac pacemakers and associated devices	F24	Insufficient Health Impact Information	A0708	Power Problem		
					CT2317	Transmitters/Receivers/Transceivers	F24	Insufficient Health Impact Information	A0708	Power Problem		
					CT233	Surgical	F24	Insufficient Health Impact Information	A0708	Power Problem		
					CT2406	Implantable	F24	Insufficient Health Impact Information	A0708	Power Problem		
					CT2407	Active implantable	F24	Insufficient Health Impact Information	A0708	Power Problem		

					CT255	Transducers and associated devices	F24	Insufficient Health Impact Information	A0708	Power Problem		
					CT2726	Physiological monitoring/measuring devices	F24	Insufficient Health Impact Information	A0708	Power Problem		
					CT2727	Active implant devices/ancillaries	F24	Insufficient Health Impact Information	A0708	Power Problem		
					CT2750	Implantable cardiac electrical stimulators and associated devices	F24	Insufficient Health Impact Information	A0708	Power Problem		
					CT2885	Physiological monitoring/measuring system transmitters/receivers/transceivers	F24	Insufficient Health Impact Information	A0708	Power Problem		
					CT2985	Active	F24	Insufficient Health Impact Information	A0708	Power Problem		
					CT3001	Implantable metal	F24	Insufficient Health Impact Information	A0708	Power Problem		
					CT3002	Implantable synthetic polymer	F24	Insufficient Health Impact Information	A0708	Power Problem		
					CT3006	Implantable inorganic material	F24	Insufficient Health Impact Information	A0708	Power Problem		
					CT314	Electrically powered	F24	Insufficient Health Impact Information	A0708	Power Problem		
					CT321	Long-term surgical invasive	F24	Insufficient Health Impact Information	A0708	Power Problem		
					CT384	Sensors and associated devices	F24	Insufficient Health Impact Information	A0708	Power Problem		
					CT449	Generators and associated devices	F24	Insufficient Health Impact Information	A0708	Power Problem		
					CT981	Single use	F24	Insufficient Health Impact Information	A0708	Power Problem		
					CT983	Surgical invasive	F24	Insufficient Health Impact Information	A0708	Power Problem		
					CT994	Sensors	F24	Insufficient Health Impact Information	A0708	Power Problem		
12369	2023	All other reportable incidents	47270	Cardiac resynchronization therapy implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	04/05/2023	04/05/2023
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	04/05/2023	04/05/2023
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	04/05/2023	04/05/2023
12370	2023	All other reportable incidents	35852	Single-chamber implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem	09/05/2023	09/05/2023
									A070504	Premature Discharge of Battery	09/05/2023	09/05/2023
							F08	Hospitalization or Prolonged Hospitalization	A0705	Battery Problem	09/05/2023	09/05/2023
									A070504	Premature Discharge of Battery	09/05/2023	09/05/2023
							F1905	Device Revision or Replacement	A0705	Battery Problem	09/05/2023	09/05/2023
									A070504	Premature Discharge of Battery	09/05/2023	09/05/2023
12371	2023		47265	Dual-chamber implantable pacemaker, rate-responsive	CT141	Electrodes/Leads and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem		
							F1901	Additional Surgery	A24	No Apparent Device Problem		
					CT145	Cardiovascular devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem		
							F1901	Additional Surgery	A24	No Apparent Device Problem		
					CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem		
							F1901	Additional Surgery	A24	No Apparent Device Problem		
					CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem		
							F1901	Additional Surgery	A24	No Apparent Device Problem		
					CT233	Surgical	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem		
							F1901	Additional Surgery	A24	No Apparent Device Problem		
					CT2406	Implantable	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem		
							F1901	Additional Surgery	A24	No Apparent Device Problem		
					CT2407	Active implantable	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem		
							F1901	Additional Surgery	A24	No Apparent Device Problem		
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem		
							F1901	Additional Surgery	A24	No Apparent Device Problem		
					CT2727	Active implant devices/ancillaries	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem		
							F1901	Additional Surgery	A24	No Apparent Device Problem		
					CT2750	Implantable cardiac electrical stimulators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem		
							F1901	Additional Surgery	A24	No Apparent Device Problem		
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem		
							F1901	Additional Surgery	A24	No Apparent Device Problem		
					CT2985	Active	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem		
							F1901	Additional Surgery	A24	No Apparent Device Problem		
					CT3001	Implantable metal	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem		
							F1901	Additional Surgery	A24	No Apparent Device Problem		
					CT3002	Implantable synthetic polymer	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem		
							F1901	Additional Surgery	A24	No Apparent Device Problem		
					CT3006	Implantable inorganic material	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem		
							F1901	Additional Surgery	A24	No Apparent Device Problem		
					CT314	Electrically powered	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem		
							F1901	Additional Surgery	A24	No Apparent Device Problem		
					CT321	Long-term surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem		
							F1901	Additional Surgery	A24	No Apparent Device Problem		
					CT449	Generators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem		
							F1901	Additional Surgery	A24	No Apparent Device Problem		
					CT981	Single use	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem		
							F1901	Additional Surgery	A24	No Apparent Device Problem		
					CT983	Surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem		
							F1901	Additional Surgery	A24	No Apparent Device Problem		
12372	2023	All other reportable incidents	35852	Single-chamber implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	20/02/2023	20/02/2023
							F1901	Additional Surgery	A070504	Premature Discharge of Battery	20/02/2023	20/02/2023
12373	2023	All other reportable incidents	47265	Dual-chamber implantable pacemaker, rate-responsive	CT141	Electrodes/Leads and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	10/02/2023	10/02/2023
									A071102	Failure to Interrogate	10/02/2023	10/02/2023
							F1901	Additional Surgery	A070504	Premature Discharge of Battery	10/02/2023	10/02/2023
									A071102	Failure to Interrogate	10/02/2023	10/02/2023
					CT145	Cardiovascular devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	10/02/2023	10/02/2023
									A071102	Failure to Interrogate	10/02/2023	10/02/2023
							F1901	Additional Surgery	A070504	Premature Discharge of Battery	10/02/2023	10/02/2023
									A071102	Failure to Interrogate	10/02/2023	10/02/2023
					CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	10/02/2023	10/02/2023
									A071102	Failure to Interrogate	10/02/2023	10/02/2023
							F1901	Additional Surgery	A070504	Premature Discharge of Battery	10/02/2023	10/02/2023
									A071102	Failure to Interrogate	10/02/2023	10/02/2023
					CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	10/02/2023	10/02/2023
									A071102	Failure to Interrogate	10/02/2023	10/02/2023
							F1901	Additional Surgery	A070504	Premature Discharge of Battery	10/02/2023	10/02/2023
									A071102	Failure to Interrogate	10/02/2023	10/02/2023
					CT233	Surgical	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	10/02/2023	10/02/2023
									A071102	Failure to Interrogate	10/02/2023	10/02/2023
							F1901	Additional Surgery	A070504	Premature Discharge of Battery	10/02/2023	10/02/2023

									A071102	Failure to Interrogate	10/02/2023	10/02/2023
					CT2406	Implantable	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	10/02/2023	10/02/2023
									A071102	Failure to Interrogate	10/02/2023	10/02/2023
							F1901	Additional Surgery	A070504	Premature Discharge of Battery	10/02/2023	10/02/2023
									A071102	Failure to Interrogate	10/02/2023	10/02/2023
					CT2407	Active implantable	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	10/02/2023	10/02/2023
									A071102	Failure to Interrogate	10/02/2023	10/02/2023
							F1901	Additional Surgery	A070504	Premature Discharge of Battery	10/02/2023	10/02/2023
									A071102	Failure to Interrogate	10/02/2023	10/02/2023
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	10/02/2023	10/02/2023
									A071102	Failure to Interrogate	10/02/2023	10/02/2023
							F1901	Additional Surgery	A070504	Premature Discharge of Battery	10/02/2023	10/02/2023
									A071102	Failure to Interrogate	10/02/2023	10/02/2023
					CT2727	Active implant devices/ancillaries	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	10/02/2023	10/02/2023
									A071102	Failure to Interrogate	10/02/2023	10/02/2023
							F1901	Additional Surgery	A070504	Premature Discharge of Battery	10/02/2023	10/02/2023
									A071102	Failure to Interrogate	10/02/2023	10/02/2023
					CT2750	Implantable cardiac electrical stimulators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	10/02/2023	10/02/2023
									A071102	Failure to Interrogate	10/02/2023	10/02/2023
							F1901	Additional Surgery	A070504	Premature Discharge of Battery	10/02/2023	10/02/2023
									A071102	Failure to Interrogate	10/02/2023	10/02/2023
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	10/02/2023	10/02/2023
									A071102	Failure to Interrogate	10/02/2023	10/02/2023
							F1901	Additional Surgery	A070504	Premature Discharge of Battery	10/02/2023	10/02/2023
									A071102	Failure to Interrogate	10/02/2023	10/02/2023
					CT2985	Active	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	10/02/2023	10/02/2023
									A071102	Failure to Interrogate	10/02/2023	10/02/2023
							F1901	Additional Surgery	A070504	Premature Discharge of Battery	10/02/2023	10/02/2023
									A071102	Failure to Interrogate	10/02/2023	10/02/2023
					CT3001	Implantable metal	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	10/02/2023	10/02/2023
									A071102	Failure to Interrogate	10/02/2023	10/02/2023
							F1901	Additional Surgery	A070504	Premature Discharge of Battery	10/02/2023	10/02/2023
									A071102	Failure to Interrogate	10/02/2023	10/02/2023
					CT3002	Implantable synthetic polymer	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	10/02/2023	10/02/2023
									A071102	Failure to Interrogate	10/02/2023	10/02/2023
							F1901	Additional Surgery	A070504	Premature Discharge of Battery	10/02/2023	10/02/2023
									A071102	Failure to Interrogate	10/02/2023	10/02/2023
					CT3006	Implantable inorganic material	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	10/02/2023	10/02/2023
									A071102	Failure to Interrogate	10/02/2023	10/02/2023
							F1901	Additional Surgery	A070504	Premature Discharge of Battery	10/02/2023	10/02/2023
									A071102	Failure to Interrogate	10/02/2023	10/02/2023
					CT314	Electrically powered	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	10/02/2023	10/02/2023
									A071102	Failure to Interrogate	10/02/2023	10/02/2023
							F1901	Additional Surgery	A070504	Premature Discharge of Battery	10/02/2023	10/02/2023
									A071102	Failure to Interrogate	10/02/2023	10/02/2023
					CT321	Long-term surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	10/02/2023	10/02/2023
									A071102	Failure to Interrogate	10/02/2023	10/02/2023
							F1901	Additional Surgery	A070504	Premature Discharge of Battery	10/02/2023	10/02/2023
									A071102	Failure to Interrogate	10/02/2023	10/02/2023
					CT449	Generators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	10/02/2023	10/02/2023
									A071102	Failure to Interrogate	10/02/2023	10/02/2023
							F1901	Additional Surgery	A070504	Premature Discharge of Battery	10/02/2023	10/02/2023
									A071102	Failure to Interrogate	10/02/2023	10/02/2023
					CT981	Single use	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	10/02/2023	10/02/2023
									A071102	Failure to Interrogate	10/02/2023	10/02/2023
							F1901	Additional Surgery	A070504	Premature Discharge of Battery	10/02/2023	10/02/2023
									A071102	Failure to Interrogate	10/02/2023	10/02/2023
					CT983	Surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	10/02/2023	10/02/2023
									A071102	Failure to Interrogate	10/02/2023	10/02/2023
							F1901	Additional Surgery	A070504	Premature Discharge of Battery	10/02/2023	10/02/2023
									A071102	Failure to Interrogate	10/02/2023	10/02/2023
12374	2023	All other reportable incidents	35852	Single-chamber implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E2104	Electric Shock	A0709	Device Sensing Problem		
									A070909	Oversensing		
									A071302	Inappropriate/Inadequate Shock/Stimulation		
									A1301	Failure to Read Input Signal		
							F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A0709	Device Sensing Problem		
									A070909	Oversensing		
									A071302	Inappropriate/Inadequate Shock/Stimulation		
									A1301	Failure to Read Input Signal		
							F2203	Imaging Required	A0709	Device Sensing Problem		
									A070909	Oversensing		
									A071302	Inappropriate/Inadequate Shock/Stimulation		
									A1301	Failure to Read Input Signal		
12375	2023	All other reportable incidents	47270	Cardiac resynchronization therapy implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A070909	Oversensing		
							F26	No Health Consequences or Impact	A070909	Oversensing		
12376	2023	All other reportable incidents	47265	Dual-chamber implantable pacemaker, rate-responsive	CT141	Electrodes/Leads and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070909	Oversensing		
									A072202	Low Impedance		
									A090801	Signal Artifact/Noise		
							F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A070909	Oversensing		
									A072202	Low Impedance		
									A090801	Signal Artifact/Noise		
							F2203	Imaging Required	A070909	Oversensing		
									A072202	Low Impedance		
									A090801	Signal Artifact/Noise		
					CT145	Cardiovascular devices	E2403	No Clinical Signs, Symptoms or Conditions	A070909	Oversensing		
									A072202	Low Impedance		
									A090801	Signal Artifact/Noise		
							F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A070909	Oversensing		

									A072202	Low Impedance		
									A090801	Signal Artifact/Noise		
							F2203	Imaging Required	A070909	Oversensing		
									A072202	Low Impedance		
									A090801	Signal Artifact/Noise		
					CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070909	Oversensing		
									A072202	Low Impedance		
									A090801	Signal Artifact/Noise		
							F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A070909	Oversensing		
									A072202	Low Impedance		
									A090801	Signal Artifact/Noise		
							F2203	Imaging Required	A070909	Oversensing		
									A072202	Low Impedance		
									A090801	Signal Artifact/Noise		
					CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A070909	Oversensing		
									A072202	Low Impedance		
									A090801	Signal Artifact/Noise		
							F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A070909	Oversensing		
									A072202	Low Impedance		
									A090801	Signal Artifact/Noise		
							F2203	Imaging Required	A070909	Oversensing		
									A072202	Low Impedance		
									A090801	Signal Artifact/Noise		
					CT233	Surgical	E2403	No Clinical Signs, Symptoms or Conditions	A070909	Oversensing		
									A072202	Low Impedance		
									A090801	Signal Artifact/Noise		
							F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A070909	Oversensing		
									A072202	Low Impedance		
									A090801	Signal Artifact/Noise		
							F2203	Imaging Required	A070909	Oversensing		
									A072202	Low Impedance		
									A090801	Signal Artifact/Noise		
					CT2406	Implantable	E2403	No Clinical Signs, Symptoms or Conditions	A070909	Oversensing		
									A072202	Low Impedance		
									A090801	Signal Artifact/Noise		
							F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A070909	Oversensing		
									A072202	Low Impedance		
									A090801	Signal Artifact/Noise		
							F2203	Imaging Required	A070909	Oversensing		
									A072202	Low Impedance		
									A090801	Signal Artifact/Noise		
					CT2407	Active implantable	E2403	No Clinical Signs, Symptoms or Conditions	A070909	Oversensing		
									A072202	Low Impedance		
									A090801	Signal Artifact/Noise		
							F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A070909	Oversensing		
									A072202	Low Impedance		
									A090801	Signal Artifact/Noise		
							F2203	Imaging Required	A070909	Oversensing		
									A072202	Low Impedance		
									A090801	Signal Artifact/Noise		
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A070909	Oversensing		
									A072202	Low Impedance		
									A090801	Signal Artifact/Noise		
							F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A070909	Oversensing		
									A072202	Low Impedance		
									A090801	Signal Artifact/Noise		
							F2203	Imaging Required	A070909	Oversensing		
									A072202	Low Impedance		
									A090801	Signal Artifact/Noise		
					CT2727	Active implant devices/ancillaries	E2403	No Clinical Signs, Symptoms or Conditions	A070909	Oversensing		
									A072202	Low Impedance		
									A090801	Signal Artifact/Noise		
							F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A070909	Oversensing		
									A072202	Low Impedance		
									A090801	Signal Artifact/Noise		
							F2203	Imaging Required	A070909	Oversensing		
									A072202	Low Impedance		
									A090801	Signal Artifact/Noise		
					CT2750	Implantable cardiac electrical stimulators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070909	Oversensing		
									A072202	Low Impedance		
									A090801	Signal Artifact/Noise		
							F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A070909	Oversensing		
									A072202	Low Impedance		
									A090801	Signal Artifact/Noise		
							F2203	Imaging Required	A070909	Oversensing		
									A072202	Low Impedance		
									A090801	Signal Artifact/Noise		
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A070909	Oversensing		
									A072202	Low Impedance		
									A090801	Signal Artifact/Noise		
							F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A070909	Oversensing		
									A072202	Low Impedance		
									A090801	Signal Artifact/Noise		
							F2203	Imaging Required	A070909	Oversensing		
									A072202	Low Impedance		
									A090801	Signal Artifact/Noise		
					CT2985	Active	E2403	No Clinical Signs, Symptoms or Conditions	A070909	Oversensing		
									A072202	Low Impedance		

									A090801	Signal Artifact/Noise		
							F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A070909	Oversensing		
									A072202	Low Impedance		
									A090801	Signal Artifact/Noise		
							F2203	Imaging Required	A070909	Oversensing		
									A072202	Low Impedance		
									A090801	Signal Artifact/Noise		
					CT3001	Implantable metal	E2403	No Clinical Signs, Symptoms or Conditions	A070909	Oversensing		
									A072202	Low Impedance		
									A090801	Signal Artifact/Noise		
							F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A070909	Oversensing		
									A072202	Low Impedance		
									A090801	Signal Artifact/Noise		
							F2203	Imaging Required	A070909	Oversensing		
									A072202	Low Impedance		
									A090801	Signal Artifact/Noise		
					CT3002	Implantable synthetic polymer	E2403	No Clinical Signs, Symptoms or Conditions	A070909	Oversensing		
									A072202	Low Impedance		
									A090801	Signal Artifact/Noise		
							F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A070909	Oversensing		
									A072202	Low Impedance		
									A090801	Signal Artifact/Noise		
							F2203	Imaging Required	A070909	Oversensing		
									A072202	Low Impedance		
									A090801	Signal Artifact/Noise		
					CT3006	Implantable inorganic material	E2403	No Clinical Signs, Symptoms or Conditions	A070909	Oversensing		
									A072202	Low Impedance		
									A090801	Signal Artifact/Noise		
							F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A070909	Oversensing		
									A072202	Low Impedance		
									A090801	Signal Artifact/Noise		
							F2203	Imaging Required	A070909	Oversensing		
									A072202	Low Impedance		
									A090801	Signal Artifact/Noise		
					CT314	Electrically powered	E2403	No Clinical Signs, Symptoms or Conditions	A070909	Oversensing		
									A072202	Low Impedance		
									A090801	Signal Artifact/Noise		
							F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A070909	Oversensing		
									A072202	Low Impedance		
									A090801	Signal Artifact/Noise		
							F2203	Imaging Required	A070909	Oversensing		
									A072202	Low Impedance		
									A090801	Signal Artifact/Noise		
					CT321	Long-term surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A070909	Oversensing		
									A072202	Low Impedance		
									A090801	Signal Artifact/Noise		
							F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A070909	Oversensing		
									A072202	Low Impedance		
									A090801	Signal Artifact/Noise		
							F2203	Imaging Required	A070909	Oversensing		
									A072202	Low Impedance		
									A090801	Signal Artifact/Noise		
					CT449	Generators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070909	Oversensing		
									A072202	Low Impedance		
									A090801	Signal Artifact/Noise		
							F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A070909	Oversensing		
									A072202	Low Impedance		
									A090801	Signal Artifact/Noise		
							F2203	Imaging Required	A070909	Oversensing		
									A072202	Low Impedance		
									A090801	Signal Artifact/Noise		
					CT981	Single use	E2403	No Clinical Signs, Symptoms or Conditions	A070909	Oversensing		
									A072202	Low Impedance		
									A090801	Signal Artifact/Noise		
							F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A070909	Oversensing		
									A072202	Low Impedance		
									A090801	Signal Artifact/Noise		
							F2203	Imaging Required	A070909	Oversensing		
									A072202	Low Impedance		
									A090801	Signal Artifact/Noise		
					CT983	Surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A070909	Oversensing		
									A072202	Low Impedance		
									A090801	Signal Artifact/Noise		
							F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A070909	Oversensing		
									A072202	Low Impedance		
									A090801	Signal Artifact/Noise		
							F2203	Imaging Required	A070909	Oversensing		
									A072202	Low Impedance		
									A090801	Signal Artifact/Noise		
12377	2023	All other reportable incidents	47270	Cardiac resynchronization therapy implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A070909	Oversensing		
									A072201	High Impedance		
									A090801	Signal Artifact/Noise		
							F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A070909	Oversensing		
									A072201	High Impedance		
									A090801	Signal Artifact/Noise		
12378	2023	All other reportable incidents	35852	Single-chamber implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem	09/11/2022	09/11/2022
							F08	Hospitalization or Prolonged Hospitalization	A0705	Battery Problem	09/11/2022	09/11/2022
							F12	Serious Injury/ Illness/ Impairment	A0705	Battery Problem	09/11/2022	09/11/2022

							F1905	Device Revision or Replacement	A0705	Battery Problem	09/11/2022	09/11/2022
12379	2023		47265	Dual-chamber implantable pacemaker, rate-responsive	CT141	Electrodes/Leads and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	20/02/2023	20/02/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	20/02/2023	20/02/2023
					CT145	Cardiovascular devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	20/02/2023	20/02/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	20/02/2023	20/02/2023
					CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	20/02/2023	20/02/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	20/02/2023	20/02/2023
					CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	20/02/2023	20/02/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	20/02/2023	20/02/2023
					CT233	Surgical	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	20/02/2023	20/02/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	20/02/2023	20/02/2023
					CT2406	Implantable	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	20/02/2023	20/02/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	20/02/2023	20/02/2023
					CT2407	Active implantable	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	20/02/2023	20/02/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	20/02/2023	20/02/2023
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	20/02/2023	20/02/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	20/02/2023	20/02/2023
					CT2727	Active implant devices/ancillaries	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	20/02/2023	20/02/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	20/02/2023	20/02/2023
					CT2750	Implantable cardiac electrical stimulators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	20/02/2023	20/02/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	20/02/2023	20/02/2023
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	20/02/2023	20/02/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	20/02/2023	20/02/2023
					CT2985	Active	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	20/02/2023	20/02/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	20/02/2023	20/02/2023
					CT3001	Implantable metal	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	20/02/2023	20/02/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	20/02/2023	20/02/2023
					CT3002	Implantable synthetic polymer	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	20/02/2023	20/02/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	20/02/2023	20/02/2023
					CT3006	Implantable inorganic material	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	20/02/2023	20/02/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	20/02/2023	20/02/2023
					CT314	Electrically powered	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	20/02/2023	20/02/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	20/02/2023	20/02/2023
					CT321	Long-term surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	20/02/2023	20/02/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	20/02/2023	20/02/2023
					CT449	Generators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	20/02/2023	20/02/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	20/02/2023	20/02/2023
					CT981	Single use	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	20/02/2023	20/02/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	20/02/2023	20/02/2023
					CT983	Surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	20/02/2023	20/02/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	20/02/2023	20/02/2023
12380	2023		47265	Dual-chamber implantable pacemaker, rate-responsive	CT141	Electrodes/Leads and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	14/02/2023	14/02/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	14/02/2023	14/02/2023
					CT145	Cardiovascular devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	14/02/2023	14/02/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	14/02/2023	14/02/2023
					CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	14/02/2023	14/02/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	14/02/2023	14/02/2023
					CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	14/02/2023	14/02/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	14/02/2023	14/02/2023
					CT233	Surgical	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	14/02/2023	14/02/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	14/02/2023	14/02/2023
					CT2406	Implantable	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	14/02/2023	14/02/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	14/02/2023	14/02/2023
					CT2407	Active implantable	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	14/02/2023	14/02/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	14/02/2023	14/02/2023
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	14/02/2023	14/02/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	14/02/2023	14/02/2023
					CT2727	Active implant devices/ancillaries	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	14/02/2023	14/02/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	14/02/2023	14/02/2023
					CT2750	Implantable cardiac electrical stimulators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	14/02/2023	14/02/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	14/02/2023	14/02/2023
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	14/02/2023	14/02/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	14/02/2023	14/02/2023
					CT2985	Active	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	14/02/2023	14/02/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	14/02/2023	14/02/2023
					CT3001	Implantable metal	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	14/02/2023	14/02/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	14/02/2023	14/02/2023
					CT3002	Implantable synthetic polymer	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	14/02/2023	14/02/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	14/02/2023	14/02/2023
					CT3006	Implantable inorganic material	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	14/02/2023	14/02/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	14/02/2023	14/02/2023
					CT314	Electrically powered	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	14/02/2023	14/02/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	14/02/2023	14/02/2023
					CT321	Long-term surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	14/02/2023	14/02/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	14/02/2023	14/02/2023
					CT449	Generators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	14/02/2023	14/02/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	14/02/2023	14/02/2023
					CT981	Single use	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	14/02/2023	14/02/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	14/02/2023	14/02/2023
					CT983	Surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	14/02/2023	14/02/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	14/02/2023	14/02/2023
12381	2023		47265	Dual-chamber implantable pacemaker, rate-responsive	CT141	Electrodes/Leads and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	12/02/2023	12/02/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	12/02/2023	12/02/2023
					CT145	Cardiovascular devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	12/02/2023	12/02/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	12/02/2023	12/02/2023
					CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	12/02/2023	12/02/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	12/02/2023	12/02/2023
					CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	12/02/2023	12/02/2023

							F1901	Additional Surgery	A24	No Apparent Device Problem	12/02/2023	12/02/2023
					CT233	Surgical	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	12/02/2023	12/02/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	12/02/2023	12/02/2023
					CT2406	Implantable	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	12/02/2023	12/02/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	12/02/2023	12/02/2023
					CT2407	Active implantable	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	12/02/2023	12/02/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	12/02/2023	12/02/2023
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	12/02/2023	12/02/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	12/02/2023	12/02/2023
					CT2727	Active implant devices/ancillaries	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	12/02/2023	12/02/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	12/02/2023	12/02/2023
					CT2750	Implantable cardiac electrical stimulators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	12/02/2023	12/02/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	12/02/2023	12/02/2023
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	12/02/2023	12/02/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	12/02/2023	12/02/2023
					CT2985	Active	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	12/02/2023	12/02/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	12/02/2023	12/02/2023
					CT3001	Implantable metal	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	12/02/2023	12/02/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	12/02/2023	12/02/2023
					CT3002	Implantable synthetic polymer	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	12/02/2023	12/02/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	12/02/2023	12/02/2023
					CT3006	Implantable inorganic material	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	12/02/2023	12/02/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	12/02/2023	12/02/2023
					CT314	Electrically powered	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	12/02/2023	12/02/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	12/02/2023	12/02/2023
					CT321	Long-term surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	12/02/2023	12/02/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	12/02/2023	12/02/2023
					CT449	Generators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	12/02/2023	12/02/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	12/02/2023	12/02/2023
					CT981	Single use	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	12/02/2023	12/02/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	12/02/2023	12/02/2023
					CT983	Surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	12/02/2023	12/02/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	12/02/2023	12/02/2023
12382	2023		47265	Dual-chamber implantable pacemaker, rate-responsive	CT141	Electrodes/Leads and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	21/02/2023	21/02/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	21/02/2023	21/02/2023
					CT145	Cardiovascular devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	21/02/2023	21/02/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	21/02/2023	21/02/2023
					CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	21/02/2023	21/02/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	21/02/2023	21/02/2023
					CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	21/02/2023	21/02/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	21/02/2023	21/02/2023
					CT233	Surgical	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	21/02/2023	21/02/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	21/02/2023	21/02/2023
					CT2406	Implantable	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	21/02/2023	21/02/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	21/02/2023	21/02/2023
					CT2407	Active implantable	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	21/02/2023	21/02/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	21/02/2023	21/02/2023
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	21/02/2023	21/02/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	21/02/2023	21/02/2023
					CT2727	Active implant devices/ancillaries	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	21/02/2023	21/02/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	21/02/2023	21/02/2023
					CT2750	Implantable cardiac electrical stimulators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	21/02/2023	21/02/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	21/02/2023	21/02/2023
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	21/02/2023	21/02/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	21/02/2023	21/02/2023
					CT2985	Active	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	21/02/2023	21/02/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	21/02/2023	21/02/2023
					CT3001	Implantable metal	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	21/02/2023	21/02/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	21/02/2023	21/02/2023
					CT3002	Implantable synthetic polymer	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	21/02/2023	21/02/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	21/02/2023	21/02/2023
					CT3006	Implantable inorganic material	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	21/02/2023	21/02/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	21/02/2023	21/02/2023
					CT314	Electrically powered	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	21/02/2023	21/02/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	21/02/2023	21/02/2023
					CT321	Long-term surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	21/02/2023	21/02/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	21/02/2023	21/02/2023
					CT449	Generators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	21/02/2023	21/02/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	21/02/2023	21/02/2023
					CT981	Single use	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	21/02/2023	21/02/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	21/02/2023	21/02/2023
					CT983	Surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	21/02/2023	21/02/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	21/02/2023	21/02/2023
12383	2023	All other reportable incidents	47265	Dual-chamber implantable pacemaker, rate-responsive	CT141	Electrodes/Leads and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	08/02/2023	08/02/2023
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	08/02/2023	08/02/2023
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	08/02/2023	08/02/2023
					CT145	Cardiovascular devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	08/02/2023	08/02/2023
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	08/02/2023	08/02/2023
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	08/02/2023	08/02/2023
					CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	08/02/2023	08/02/2023
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	08/02/2023	08/02/2023
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	08/02/2023	08/02/2023
					CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	08/02/2023	08/02/2023
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	08/02/2023	08/02/2023
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	08/02/2023	08/02/2023
					CT233	Surgical	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	08/02/2023	08/02/2023
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	08/02/2023	08/02/2023
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	08/02/2023	08/02/2023

					CT2406	Implantable	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	08/02/2023	08/02/2023
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	08/02/2023	08/02/2023
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	08/02/2023	08/02/2023
					CT2407	Active implantable	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	08/02/2023	08/02/2023
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	08/02/2023	08/02/2023
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	08/02/2023	08/02/2023
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	08/02/2023	08/02/2023
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	08/02/2023	08/02/2023
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	08/02/2023	08/02/2023
					CT2727	Active implant devices/ancillaries	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	08/02/2023	08/02/2023
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	08/02/2023	08/02/2023
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	08/02/2023	08/02/2023
					CT2750	Implantable cardiac electrical stimulators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	08/02/2023	08/02/2023
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	08/02/2023	08/02/2023
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	08/02/2023	08/02/2023
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	08/02/2023	08/02/2023
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	08/02/2023	08/02/2023
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	08/02/2023	08/02/2023
					CT2985	Active	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	08/02/2023	08/02/2023
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	08/02/2023	08/02/2023
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	08/02/2023	08/02/2023
					CT3001	Implantable metal	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	08/02/2023	08/02/2023
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	08/02/2023	08/02/2023
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	08/02/2023	08/02/2023
					CT3002	Implantable synthetic polymer	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	08/02/2023	08/02/2023
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	08/02/2023	08/02/2023
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	08/02/2023	08/02/2023
					CT3006	Implantable inorganic material	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	08/02/2023	08/02/2023
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	08/02/2023	08/02/2023
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	08/02/2023	08/02/2023
					CT314	Electrically powered	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	08/02/2023	08/02/2023
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	08/02/2023	08/02/2023
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	08/02/2023	08/02/2023
					CT321	Long-term surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	08/02/2023	08/02/2023
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	08/02/2023	08/02/2023
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	08/02/2023	08/02/2023
					CT449	Generators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	08/02/2023	08/02/2023
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	08/02/2023	08/02/2023
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	08/02/2023	08/02/2023
					CT981	Single use	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	08/02/2023	08/02/2023
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	08/02/2023	08/02/2023
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	08/02/2023	08/02/2023
					CT983	Surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	08/02/2023	08/02/2023
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	08/02/2023	08/02/2023
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	08/02/2023	08/02/2023
12384	2023		47265	Dual-chamber implantable pacemaker, rate-responsive	CT141	Electrodes/Leads and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	19/02/2023	19/02/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	19/02/2023	19/02/2023
					CT145	Cardiovascular devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	19/02/2023	19/02/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	19/02/2023	19/02/2023
					CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	19/02/2023	19/02/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	19/02/2023	19/02/2023
					CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	19/02/2023	19/02/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	19/02/2023	19/02/2023
					CT233	Surgical	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	19/02/2023	19/02/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	19/02/2023	19/02/2023
					CT2406	Implantable	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	19/02/2023	19/02/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	19/02/2023	19/02/2023
					CT2407	Active implantable	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	19/02/2023	19/02/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	19/02/2023	19/02/2023
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	19/02/2023	19/02/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	19/02/2023	19/02/2023
					CT2727	Active implant devices/ancillaries	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	19/02/2023	19/02/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	19/02/2023	19/02/2023
					CT2750	Implantable cardiac electrical stimulators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	19/02/2023	19/02/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	19/02/2023	19/02/2023
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	19/02/2023	19/02/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	19/02/2023	19/02/2023
					CT2985	Active	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	19/02/2023	19/02/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	19/02/2023	19/02/2023
					CT3001	Implantable metal	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	19/02/2023	19/02/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	19/02/2023	19/02/2023
					CT3002	Implantable synthetic polymer	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	19/02/2023	19/02/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	19/02/2023	19/02/2023
					CT3006	Implantable inorganic material	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	19/02/2023	19/02/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	19/02/2023	19/02/2023
					CT314	Electrically powered	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	19/02/2023	19/02/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	19/02/2023	19/02/2023
					CT321	Long-term surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	19/02/2023	19/02/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	19/02/2023	19/02/2023
					CT449	Generators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	19/02/2023	19/02/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	19/02/2023	19/02/2023
					CT981	Single use	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	19/02/2023	19/02/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	19/02/2023	19/02/2023
					CT983	Surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	19/02/2023	19/02/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	19/02/2023	19/02/2023
12385	2023		47265	Dual-chamber implantable pacemaker, rate-responsive	CT141	Electrodes/Leads and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	19/02/2023	19/02/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	19/02/2023	19/02/2023
					CT145	Cardiovascular devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	19/02/2023	19/02/2023

							F1901	Additional Surgery	A24	No Apparent Device Problem	19/02/2023	19/02/2023
					CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	19/02/2023	19/02/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	19/02/2023	19/02/2023
					CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	19/02/2023	19/02/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	19/02/2023	19/02/2023
					CT233	Surgical	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	19/02/2023	19/02/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	19/02/2023	19/02/2023
					CT2406	Implantable	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	19/02/2023	19/02/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	19/02/2023	19/02/2023
					CT2407	Active implantable	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	19/02/2023	19/02/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	19/02/2023	19/02/2023
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	19/02/2023	19/02/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	19/02/2023	19/02/2023
					CT2727	Active implant devices/ancillaries	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	19/02/2023	19/02/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	19/02/2023	19/02/2023
					CT2750	Implantable cardiac electrical stimulators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	19/02/2023	19/02/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	19/02/2023	19/02/2023
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	19/02/2023	19/02/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	19/02/2023	19/02/2023
					CT2985	Active	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	19/02/2023	19/02/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	19/02/2023	19/02/2023
					CT3001	Implantable metal	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	19/02/2023	19/02/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	19/02/2023	19/02/2023
					CT3002	Implantable synthetic polymer	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	19/02/2023	19/02/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	19/02/2023	19/02/2023
					CT3006	Implantable inorganic material	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	19/02/2023	19/02/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	19/02/2023	19/02/2023
					CT314	Electrically powered	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	19/02/2023	19/02/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	19/02/2023	19/02/2023
					CT321	Long-term surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	19/02/2023	19/02/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	19/02/2023	19/02/2023
					CT449	Generators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	19/02/2023	19/02/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	19/02/2023	19/02/2023
					CT981	Single use	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	19/02/2023	19/02/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	19/02/2023	19/02/2023
					CT983	Surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	19/02/2023	19/02/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	19/02/2023	19/02/2023
12386	2023		47265	Dual-chamber implantable pacemaker, rate-responsive	CT141	Electrodes/Leads and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	20/02/2023	20/02/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	20/02/2023	20/02/2023
					CT145	Cardiovascular devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	20/02/2023	20/02/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	20/02/2023	20/02/2023
					CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	20/02/2023	20/02/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	20/02/2023	20/02/2023
					CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	20/02/2023	20/02/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	20/02/2023	20/02/2023
					CT233	Surgical	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	20/02/2023	20/02/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	20/02/2023	20/02/2023
					CT2406	Implantable	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	20/02/2023	20/02/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	20/02/2023	20/02/2023
					CT2407	Active implantable	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	20/02/2023	20/02/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	20/02/2023	20/02/2023
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	20/02/2023	20/02/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	20/02/2023	20/02/2023
					CT2727	Active implant devices/ancillaries	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	20/02/2023	20/02/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	20/02/2023	20/02/2023
					CT2750	Implantable cardiac electrical stimulators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	20/02/2023	20/02/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	20/02/2023	20/02/2023
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	20/02/2023	20/02/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	20/02/2023	20/02/2023
					CT2985	Active	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	20/02/2023	20/02/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	20/02/2023	20/02/2023
					CT3001	Implantable metal	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	20/02/2023	20/02/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	20/02/2023	20/02/2023
					CT3002	Implantable synthetic polymer	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	20/02/2023	20/02/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	20/02/2023	20/02/2023
					CT3006	Implantable inorganic material	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	20/02/2023	20/02/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	20/02/2023	20/02/2023
					CT314	Electrically powered	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	20/02/2023	20/02/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	20/02/2023	20/02/2023
					CT321	Long-term surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	20/02/2023	20/02/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	20/02/2023	20/02/2023
					CT449	Generators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	20/02/2023	20/02/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	20/02/2023	20/02/2023
					CT981	Single use	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	20/02/2023	20/02/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	20/02/2023	20/02/2023
					CT983	Surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	20/02/2023	20/02/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	20/02/2023	20/02/2023
12387	2023		47265	Dual-chamber implantable pacemaker, rate-responsive	CT141	Electrodes/Leads and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	19/02/2023	19/02/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	19/02/2023	19/02/2023
					CT145	Cardiovascular devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	19/02/2023	19/02/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	19/02/2023	19/02/2023
					CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	19/02/2023	19/02/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	19/02/2023	19/02/2023
					CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	19/02/2023	19/02/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	19/02/2023	19/02/2023
					CT233	Surgical	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	19/02/2023	19/02/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	19/02/2023	19/02/2023
					CT2406	Implantable	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	19/02/2023	19/02/2023

							F1901	Additional Surgery	A24	No Apparent Device Problem	19/02/2023	19/02/2023
					CT2407	Active implantable	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	19/02/2023	19/02/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	19/02/2023	19/02/2023
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	19/02/2023	19/02/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	19/02/2023	19/02/2023
					CT2727	Active implant devices/ancillaries	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	19/02/2023	19/02/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	19/02/2023	19/02/2023
					CT2750	Implantable cardiac electrical stimulators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	19/02/2023	19/02/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	19/02/2023	19/02/2023
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	19/02/2023	19/02/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	19/02/2023	19/02/2023
					CT2985	Active	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	19/02/2023	19/02/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	19/02/2023	19/02/2023
					CT3001	Implantable metal	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	19/02/2023	19/02/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	19/02/2023	19/02/2023
					CT3002	Implantable synthetic polymer	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	19/02/2023	19/02/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	19/02/2023	19/02/2023
					CT3006	Implantable inorganic material	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	19/02/2023	19/02/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	19/02/2023	19/02/2023
					CT314	Electrically powered	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	19/02/2023	19/02/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	19/02/2023	19/02/2023
					CT321	Long-term surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	19/02/2023	19/02/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	19/02/2023	19/02/2023
					CT449	Generators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	19/02/2023	19/02/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	19/02/2023	19/02/2023
					CT981	Single use	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	19/02/2023	19/02/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	19/02/2023	19/02/2023
					CT983	Surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	19/02/2023	19/02/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	19/02/2023	19/02/2023
12388	2023		47265	Dual-chamber implantable pacemaker, rate-responsive	CT141	Electrodes/Leads and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	19/02/2023	19/02/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	19/02/2023	19/02/2023
					CT145	Cardiovascular devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	19/02/2023	19/02/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	19/02/2023	19/02/2023
					CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	19/02/2023	19/02/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	19/02/2023	19/02/2023
					CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	19/02/2023	19/02/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	19/02/2023	19/02/2023
					CT233	Surgical	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	19/02/2023	19/02/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	19/02/2023	19/02/2023
					CT2406	Implantable	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	19/02/2023	19/02/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	19/02/2023	19/02/2023
					CT2407	Active implantable	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	19/02/2023	19/02/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	19/02/2023	19/02/2023
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	19/02/2023	19/02/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	19/02/2023	19/02/2023
					CT2727	Active implant devices/ancillaries	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	19/02/2023	19/02/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	19/02/2023	19/02/2023
					CT2750	Implantable cardiac electrical stimulators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	19/02/2023	19/02/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	19/02/2023	19/02/2023
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	19/02/2023	19/02/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	19/02/2023	19/02/2023
					CT2985	Active	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	19/02/2023	19/02/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	19/02/2023	19/02/2023
					CT3001	Implantable metal	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	19/02/2023	19/02/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	19/02/2023	19/02/2023
					CT3002	Implantable synthetic polymer	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	19/02/2023	19/02/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	19/02/2023	19/02/2023
					CT3006	Implantable inorganic material	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	19/02/2023	19/02/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	19/02/2023	19/02/2023
					CT314	Electrically powered	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	19/02/2023	19/02/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	19/02/2023	19/02/2023
					CT321	Long-term surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	19/02/2023	19/02/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	19/02/2023	19/02/2023
					CT449	Generators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	19/02/2023	19/02/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	19/02/2023	19/02/2023
					CT981	Single use	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	19/02/2023	19/02/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	19/02/2023	19/02/2023
					CT983	Surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	19/02/2023	19/02/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	19/02/2023	19/02/2023
12389	2023	All other reportable incidents	17882	Physiologic-monitoring defibrillation system	CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A071301	Failure to Deliver Shock/Stimulation		
									A0719	Unexpected Shutdown		
							F26	No Health Consequences or Impact	A071301	Failure to Deliver Shock/Stimulation		
									A0719	Unexpected Shutdown		
12390	2023	All other reportable incidents	47265	Dual-chamber implantable pacemaker, rate-responsive	CT141	Electrodes/Leads and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070909	Oversensing		
									A0712	Pacing Problem		
									A072202	Low Impedance		
									A090801	Signal Artifact/Noise		
									A1301	Failure to Read Input Signal		
							F26	No Health Consequences or Impact	A070909	Oversensing		
									A0712	Pacing Problem		
									A072202	Low Impedance		
									A090801	Signal Artifact/Noise		
									A1301	Failure to Read Input Signal		
					CT145	Cardiovascular devices	E2403	No Clinical Signs, Symptoms or Conditions	A070909	Oversensing		
									A0712	Pacing Problem		
									A072202	Low Impedance		
									A090801	Signal Artifact/Noise		
									A1301	Failure to Read Input Signal		

							F26	No Health Consequences or Impact	A070909	Oversensing		
									A0712	Pacing Problem		
									A072202	Low Impedance		
									A090801	Signal Artifact/Noise		
									A1301	Failure to Read Input Signal		
					CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070909	Oversensing		
									A0712	Pacing Problem		
									A072202	Low Impedance		
									A090801	Signal Artifact/Noise		
									A1301	Failure to Read Input Signal		
							F26	No Health Consequences or Impact	A070909	Oversensing		
									A0712	Pacing Problem		
									A072202	Low Impedance		
									A090801	Signal Artifact/Noise		
									A1301	Failure to Read Input Signal		
					CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A070909	Oversensing		
									A0712	Pacing Problem		
									A072202	Low Impedance		
									A090801	Signal Artifact/Noise		
									A1301	Failure to Read Input Signal		
							F26	No Health Consequences or Impact	A070909	Oversensing		
									A0712	Pacing Problem		
									A072202	Low Impedance		
									A090801	Signal Artifact/Noise		
									A1301	Failure to Read Input Signal		
					CT233	Surgical	E2403	No Clinical Signs, Symptoms or Conditions	A070909	Oversensing		
									A0712	Pacing Problem		
									A072202	Low Impedance		
									A090801	Signal Artifact/Noise		
									A1301	Failure to Read Input Signal		
							F26	No Health Consequences or Impact	A070909	Oversensing		
									A0712	Pacing Problem		
									A072202	Low Impedance		
									A090801	Signal Artifact/Noise		
									A1301	Failure to Read Input Signal		
					CT2406	Implantable	E2403	No Clinical Signs, Symptoms or Conditions	A070909	Oversensing		
									A0712	Pacing Problem		
									A072202	Low Impedance		
									A090801	Signal Artifact/Noise		
									A1301	Failure to Read Input Signal		
							F26	No Health Consequences or Impact	A070909	Oversensing		
									A0712	Pacing Problem		
									A072202	Low Impedance		
									A090801	Signal Artifact/Noise		
									A1301	Failure to Read Input Signal		
					CT2407	Active implantable	E2403	No Clinical Signs, Symptoms or Conditions	A070909	Oversensing		
									A0712	Pacing Problem		
									A072202	Low Impedance		
									A090801	Signal Artifact/Noise		
									A1301	Failure to Read Input Signal		
							F26	No Health Consequences or Impact	A070909	Oversensing		
									A0712	Pacing Problem		
									A072202	Low Impedance		
									A090801	Signal Artifact/Noise		
									A1301	Failure to Read Input Signal		
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A070909	Oversensing		
									A0712	Pacing Problem		
									A072202	Low Impedance		
									A090801	Signal Artifact/Noise		
									A1301	Failure to Read Input Signal		
							F26	No Health Consequences or Impact	A070909	Oversensing		
									A0712	Pacing Problem		
									A072202	Low Impedance		
									A090801	Signal Artifact/Noise		
									A1301	Failure to Read Input Signal		
					CT2727	Active implant devices/ancillaries	E2403	No Clinical Signs, Symptoms or Conditions	A070909	Oversensing		
									A0712	Pacing Problem		
									A072202	Low Impedance		
									A090801	Signal Artifact/Noise		
									A1301	Failure to Read Input Signal		
							F26	No Health Consequences or Impact	A070909	Oversensing		
									A0712	Pacing Problem		
									A072202	Low Impedance		
									A090801	Signal Artifact/Noise		
									A1301	Failure to Read Input Signal		
					CT2750	Implantable cardiac electrical stimulators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070909	Oversensing		
									A0712	Pacing Problem		
									A072202	Low Impedance		
									A090801	Signal Artifact/Noise		
									A1301	Failure to Read Input Signal		
							F26	No Health Consequences or Impact	A070909	Oversensing		
									A0712	Pacing Problem		
									A072202	Low Impedance		
									A090801	Signal Artifact/Noise		
									A1301	Failure to Read Input Signal		
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A070909	Oversensing		
									A0712	Pacing Problem		
									A072202	Low Impedance		

									A090801	Signal Artifact/Noise		
									A1301	Failure to Read Input Signal		
							F26	No Health Consequences or Impact	A070909	Oversensing		
									A0712	Pacing Problem		
									A072202	Low Impedance		
									A090801	Signal Artifact/Noise		
									A1301	Failure to Read Input Signal		
					CT2985	Active	E2403	No Clinical Signs, Symptoms or Conditions	A070909	Oversensing		
									A0712	Pacing Problem		
									A072202	Low Impedance		
									A090801	Signal Artifact/Noise		
									A1301	Failure to Read Input Signal		
							F26	No Health Consequences or Impact	A070909	Oversensing		
									A0712	Pacing Problem		
									A072202	Low Impedance		
									A090801	Signal Artifact/Noise		
									A1301	Failure to Read Input Signal		
					CT3001	Implantable metal	E2403	No Clinical Signs, Symptoms or Conditions	A070909	Oversensing		
									A0712	Pacing Problem		
									A072202	Low Impedance		
									A090801	Signal Artifact/Noise		
									A1301	Failure to Read Input Signal		
							F26	No Health Consequences or Impact	A070909	Oversensing		
									A0712	Pacing Problem		
									A072202	Low Impedance		
									A090801	Signal Artifact/Noise		
									A1301	Failure to Read Input Signal		
					CT3002	Implantable synthetic polymer	E2403	No Clinical Signs, Symptoms or Conditions	A070909	Oversensing		
									A0712	Pacing Problem		
									A072202	Low Impedance		
									A090801	Signal Artifact/Noise		
									A1301	Failure to Read Input Signal		
							F26	No Health Consequences or Impact	A070909	Oversensing		
									A0712	Pacing Problem		
									A072202	Low Impedance		
									A090801	Signal Artifact/Noise		
									A1301	Failure to Read Input Signal		
					CT3006	Implantable inorganic material	E2403	No Clinical Signs, Symptoms or Conditions	A070909	Oversensing		
									A0712	Pacing Problem		
									A072202	Low Impedance		
									A090801	Signal Artifact/Noise		
									A1301	Failure to Read Input Signal		
							F26	No Health Consequences or Impact	A070909	Oversensing		
									A0712	Pacing Problem		
									A072202	Low Impedance		
									A090801	Signal Artifact/Noise		
									A1301	Failure to Read Input Signal		
					CT314	Electrically powered	E2403	No Clinical Signs, Symptoms or Conditions	A070909	Oversensing		
									A0712	Pacing Problem		
									A072202	Low Impedance		
									A090801	Signal Artifact/Noise		
									A1301	Failure to Read Input Signal		
							F26	No Health Consequences or Impact	A070909	Oversensing		
									A0712	Pacing Problem		
									A072202	Low Impedance		
									A090801	Signal Artifact/Noise		
									A1301	Failure to Read Input Signal		
					CT321	Long-term surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A070909	Oversensing		
									A0712	Pacing Problem		
									A072202	Low Impedance		
									A090801	Signal Artifact/Noise		
									A1301	Failure to Read Input Signal		
							F26	No Health Consequences or Impact	A070909	Oversensing		
									A0712	Pacing Problem		
									A072202	Low Impedance		
									A090801	Signal Artifact/Noise		
									A1301	Failure to Read Input Signal		
					CT449	Generators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070909	Oversensing		
									A0712	Pacing Problem		
									A072202	Low Impedance		
									A090801	Signal Artifact/Noise		
									A1301	Failure to Read Input Signal		
							F26	No Health Consequences or Impact	A070909	Oversensing		
									A0712	Pacing Problem		
									A072202	Low Impedance		
									A090801	Signal Artifact/Noise		
									A1301	Failure to Read Input Signal		
					CT981	Single use	E2403	No Clinical Signs, Symptoms or Conditions	A070909	Oversensing		
									A0712	Pacing Problem		
									A072202	Low Impedance		
									A090801	Signal Artifact/Noise		
									A1301	Failure to Read Input Signal		
							F26	No Health Consequences or Impact	A070909	Oversensing		
									A0712	Pacing Problem		
									A072202	Low Impedance		
									A090801	Signal Artifact/Noise		
									A1301	Failure to Read Input Signal		
					CT983	Surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A070909	Oversensing		

									A0712	Pacing Problem		
									A072202	Low Impedance		
									A090801	Signal Artifact/Noise		
									A1301	Failure to Read Input Signal		
							F26	No Health Consequences or Impact	A070909	Oversensing		
									A0712	Pacing Problem		
									A072202	Low Impedance		
									A090801	Signal Artifact/Noise		
									A1301	Failure to Read Input Signal		
12391	2023		47265	Dual-chamber implantable pacemaker, rate-responsive	CT141	Electrodes/Leads and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	29/01/2023	29/01/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	29/01/2023	29/01/2023
					CT145	Cardiovascular devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	29/01/2023	29/01/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	29/01/2023	29/01/2023
					CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	29/01/2023	29/01/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	29/01/2023	29/01/2023
					CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	29/01/2023	29/01/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	29/01/2023	29/01/2023
					CT233	Surgical	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	29/01/2023	29/01/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	29/01/2023	29/01/2023
					CT2406	Implantable	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	29/01/2023	29/01/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	29/01/2023	29/01/2023
					CT2407	Active implantable	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	29/01/2023	29/01/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	29/01/2023	29/01/2023
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	29/01/2023	29/01/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	29/01/2023	29/01/2023
					CT2727	Active implant devices/ancillaries	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	29/01/2023	29/01/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	29/01/2023	29/01/2023
					CT2750	Implantable cardiac electrical stimulators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	29/01/2023	29/01/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	29/01/2023	29/01/2023
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	29/01/2023	29/01/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	29/01/2023	29/01/2023
					CT2985	Active	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	29/01/2023	29/01/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	29/01/2023	29/01/2023
					CT3001	Implantable metal	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	29/01/2023	29/01/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	29/01/2023	29/01/2023
					CT3002	Implantable synthetic polymer	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	29/01/2023	29/01/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	29/01/2023	29/01/2023
					CT3006	Implantable inorganic material	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	29/01/2023	29/01/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	29/01/2023	29/01/2023
					CT314	Electrically powered	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	29/01/2023	29/01/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	29/01/2023	29/01/2023
					CT321	Long-term surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	29/01/2023	29/01/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	29/01/2023	29/01/2023
					CT449	Generators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	29/01/2023	29/01/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	29/01/2023	29/01/2023
					CT981	Single use	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	29/01/2023	29/01/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	29/01/2023	29/01/2023
					CT983	Surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	29/01/2023	29/01/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	29/01/2023	29/01/2023
12392	2023	All other reportable incidents	36241	Pacemaker/defibrillator lead	CT1684	Cardiac pacemakers and associated devices	E2104	Electric Shock	A072201	High Impedance		
							F22	Unexpected Diagnostic Intervention	A072201	High Impedance		
12393	2023	All other reportable incidents	35852	Single-chamber implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E2104	Electric Shock	A070504	Premature Discharge of Battery		
									A072201	High Impedance		
							F22	Unexpected Diagnostic Intervention	A070504	Premature Discharge of Battery		
									A072201	High Impedance		
12394	2023	All other reportable incidents	47265	Dual-chamber implantable pacemaker, rate-responsive	CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F12	Serious Injury/ Illness/ Impairment	A070504	Premature Discharge of Battery		
							F23	Unexpected Medical Intervention	A070504	Premature Discharge of Battery		
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F12	Serious Injury/ Illness/ Impairment	A070504	Premature Discharge of Battery		
							F23	Unexpected Medical Intervention	A070504	Premature Discharge of Battery		
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F12	Serious Injury/ Illness/ Impairment	A070504	Premature Discharge of Battery		
							F23	Unexpected Medical Intervention	A070504	Premature Discharge of Battery		
12395	2023	All other reportable incidents	47263	Cardiac resynchronization therapy implantable pacemaker	CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A0908	Incorrect, Inadequate or Imprecise Result or Readings	30/12/2022	30/12/2022
									A1601	Device Alarm System	30/12/2022	30/12/2022
							F08	Hospitalization or Prolonged Hospitalization	A0908	Incorrect, Inadequate or Imprecise Result or Readings	30/12/2022	30/12/2022
									A1601	Device Alarm System	30/12/2022	30/12/2022
							F1905	Device Revision or Replacement	A0908	Incorrect, Inadequate or Imprecise Result or Readings	30/12/2022	30/12/2022
									A1601	Device Alarm System	30/12/2022	30/12/2022
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A0908	Incorrect, Inadequate or Imprecise Result or Readings	30/12/2022	30/12/2022
									A1601	Device Alarm System	30/12/2022	30/12/2022
							F08	Hospitalization or Prolonged Hospitalization	A0908	Incorrect, Inadequate or Imprecise Result or Readings	30/12/2022	30/12/2022
									A1601	Device Alarm System	30/12/2022	30/12/2022
							F1905	Device Revision or Replacement	A0908	Incorrect, Inadequate or Imprecise Result or Readings	30/12/2022	30/12/2022
									A1601	Device Alarm System	30/12/2022	30/12/2022
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A0908	Incorrect, Inadequate or Imprecise Result or Readings	30/12/2022	30/12/2022
									A1601	Device Alarm System	30/12/2022	30/12/2022
							F08	Hospitalization or Prolonged Hospitalization	A0908	Incorrect, Inadequate or Imprecise Result or Readings	30/12/2022	30/12/2022
									A1601	Device Alarm System	30/12/2022	30/12/2022
							F1905	Device Revision or Replacement	A0908	Incorrect, Inadequate or Imprecise Result or Readings	30/12/2022	30/12/2022

									A1601	Device Alarm System	30/12/2022	30/12/2022
12396	2023	All other reportable incidents	47265	Dual-chamber implantable pacemaker, rate-responsive	CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
									A0908	Incorrect, Inadequate or Imprecise Result or Readings		
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery		
									A0908	Incorrect, Inadequate or Imprecise Result or Readings		
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
									A0908	Incorrect, Inadequate or Imprecise Result or Readings		
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery		
									A0908	Incorrect, Inadequate or Imprecise Result or Readings		
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
									A0908	Incorrect, Inadequate or Imprecise Result or Readings		
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery		
									A0908	Incorrect, Inadequate or Imprecise Result or Readings		
12397	2023	All other reportable incidents	35852	Single-chamber implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
12398	2023	All other reportable incidents	47265	Dual-chamber implantable pacemaker, rate-responsive	CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem	12/01/2023	12/01/2023
									A070504	Premature Discharge of Battery	12/01/2023	12/01/2023
							F26	No Health Consequences or Impact	A0705	Battery Problem	12/01/2023	12/01/2023
									A070504	Premature Discharge of Battery	12/01/2023	12/01/2023
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem	12/01/2023	12/01/2023
									A070504	Premature Discharge of Battery	12/01/2023	12/01/2023
							F26	No Health Consequences or Impact	A0705	Battery Problem	12/01/2023	12/01/2023
									A070504	Premature Discharge of Battery	12/01/2023	12/01/2023
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem	12/01/2023	12/01/2023
									A070504	Premature Discharge of Battery	12/01/2023	12/01/2023
							F26	No Health Consequences or Impact	A0705	Battery Problem	12/01/2023	12/01/2023
									A070504	Premature Discharge of Battery	12/01/2023	12/01/2023
12399	2023	All other reportable incidents	35852	Single-chamber implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	15/02/2023	15/02/2023
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	15/02/2023	15/02/2023
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	15/02/2023	15/02/2023
12400	2023	All other reportable incidents	47265	Dual-chamber implantable pacemaker, rate-responsive	CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
12401	2023	All other reportable incidents	35852	Single-chamber implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
12402	2023	All other reportable incidents	47265	Dual-chamber implantable pacemaker, rate-responsive	CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A0722	Impedance Problem		
									A072201	High Impedance		
							F26	No Health Consequences or Impact	A0722	Impedance Problem		
									A072201	High Impedance		
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A0722	Impedance Problem		
									A072201	High Impedance		
							F26	No Health Consequences or Impact	A0722	Impedance Problem		
									A072201	High Impedance		
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A0722	Impedance Problem		
									A072201	High Impedance		
							F26	No Health Consequences or Impact	A0722	Impedance Problem		
									A072201	High Impedance		
12403	2023	Death	37265	Dual-chamber implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E2401	Insufficient Information	A0709	Device Sensing Problem		
							F02	Death	A0709	Device Sensing Problem		
12404	2023	All other reportable incidents	47265	Dual-chamber implantable pacemaker, rate-responsive	CT141	Electrodes/Leads and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
					CT145	Cardiovascular devices	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
					CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
					CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
					CT233	Surgical	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
					CT2406	Implantable	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem		

							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
									A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
					CT2407	Active implantable	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
					CT2727	Active implant devices/ancillaries	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
					CT2750	Implantable cardiac electrical stimulators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
					CT2985	Active	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
					CT3001	Implantable metal	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
					CT3002	Implantable synthetic polymer	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
					CT3006	Implantable inorganic material	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
					CT314	Electrically powered	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
					CT321	Long-term surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
					CT449	Generators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
					CT981	Single use	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
					CT983	Surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
12405	2023	All other reportable incidents	17882	Physiologic-monitoring defibrillation system	CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A0719	Unexpected Shutdown		
							F27	Problem Identified During Non-Clinical Procedure	A0719	Unexpected Shutdown		
12406	2023		47265	Dual-chamber implantable pacemaker, rate-responsive	CT141	Electrodes/Leads and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	05/03/2023	05/03/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	05/03/2023	05/03/2023
					CT145	Cardiovascular devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	05/03/2023	05/03/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	05/03/2023	05/03/2023
					CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	05/03/2023	05/03/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	05/03/2023	05/03/2023
					CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	05/03/2023	05/03/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	05/03/2023	05/03/2023
					CT233	Surgical	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	05/03/2023	05/03/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	05/03/2023	05/03/2023
					CT2406	Implantable	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	05/03/2023	05/03/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	05/03/2023	05/03/2023
					CT2407	Active implantable	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	05/03/2023	05/03/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	05/03/2023	05/03/2023
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	05/03/2023	05/03/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	05/03/2023	05/03/2023
					CT2727	Active implant devices/ancillaries	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	05/03/2023	05/03/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	05/03/2023	05/03/2023
					CT2750	Implantable cardiac electrical stimulators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	05/03/2023	05/03/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	05/03/2023	05/03/2023
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	05/03/2023	05/03/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	05/03/2023	05/03/2023
					CT2985	Active	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	05/03/2023	05/03/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	05/03/2023	05/03/2023
					CT3001	Implantable metal	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	05/03/2023	05/03/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	05/03/2023	05/03/2023
					CT3002	Implantable synthetic polymer	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	05/03/2023	05/03/2023

						F1901	Additional Surgery	A24	No Apparent Device Problem	05/03/2023	05/03/2023
				CT3006	Implantable inorganic material	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	05/03/2023	05/03/2023
						F1901	Additional Surgery	A24	No Apparent Device Problem	05/03/2023	05/03/2023
				CT314	Electrically powered	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	05/03/2023	05/03/2023
						F1901	Additional Surgery	A24	No Apparent Device Problem	05/03/2023	05/03/2023
				CT321	Long-term surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	05/03/2023	05/03/2023
						F1901	Additional Surgery	A24	No Apparent Device Problem	05/03/2023	05/03/2023
				CT449	Generators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	05/03/2023	05/03/2023
						F1901	Additional Surgery	A24	No Apparent Device Problem	05/03/2023	05/03/2023
				CT981	Single use	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	05/03/2023	05/03/2023
						F1901	Additional Surgery	A24	No Apparent Device Problem	05/03/2023	05/03/2023
				CT983	Surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	05/03/2023	05/03/2023
						F1901	Additional Surgery	A24	No Apparent Device Problem	05/03/2023	05/03/2023
12407	2023	All other reportable incidents	35223	Endocardial/interventricular septal pacing lead	CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A0701	Capturing Problem		
									A070909	Oversensing		
									A072202	Low Impedance		
							F1908	Prolonged Surgery	A0701	Capturing Problem		
									A070909	Oversensing		
									A072202	Low Impedance		
12408	2023		47265	Dual-chamber implantable pacemaker, rate-responsive	CT141	Electrodes/Leads and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	19/02/2023	19/02/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	19/02/2023	19/02/2023
					CT145	Cardiovascular devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	19/02/2023	19/02/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	19/02/2023	19/02/2023
					CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	19/02/2023	19/02/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	19/02/2023	19/02/2023
					CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	19/02/2023	19/02/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	19/02/2023	19/02/2023
					CT233	Surgical	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	19/02/2023	19/02/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	19/02/2023	19/02/2023
					CT2406	Implantable	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	19/02/2023	19/02/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	19/02/2023	19/02/2023
					CT2407	Active implantable	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	19/02/2023	19/02/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	19/02/2023	19/02/2023
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	19/02/2023	19/02/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	19/02/2023	19/02/2023
					CT2727	Active implant devices/ancillaries	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	19/02/2023	19/02/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	19/02/2023	19/02/2023
					CT2750	Implantable cardiac electrical stimulators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	19/02/2023	19/02/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	19/02/2023	19/02/2023
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	19/02/2023	19/02/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	19/02/2023	19/02/2023
					CT2985	Active	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	19/02/2023	19/02/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	19/02/2023	19/02/2023
					CT3001	Implantable metal	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	19/02/2023	19/02/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	19/02/2023	19/02/2023
					CT3002	Implantable synthetic polymer	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	19/02/2023	19/02/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	19/02/2023	19/02/2023
					CT3006	Implantable inorganic material	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	19/02/2023	19/02/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	19/02/2023	19/02/2023
					CT314	Electrically powered	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	19/02/2023	19/02/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	19/02/2023	19/02/2023
					CT321	Long-term surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	19/02/2023	19/02/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	19/02/2023	19/02/2023
					CT449	Generators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	19/02/2023	19/02/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	19/02/2023	19/02/2023
					CT981	Single use	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	19/02/2023	19/02/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	19/02/2023	19/02/2023
					CT983	Surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	19/02/2023	19/02/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	19/02/2023	19/02/2023
12409	2023		47265	Dual-chamber implantable pacemaker, rate-responsive	CT141	Electrodes/Leads and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	08/02/2023	08/02/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	08/02/2023	08/02/2023
					CT145	Cardiovascular devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	08/02/2023	08/02/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	08/02/2023	08/02/2023
					CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	08/02/2023	08/02/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	08/02/2023	08/02/2023
					CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	08/02/2023	08/02/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	08/02/2023	08/02/2023
					CT233	Surgical	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	08/02/2023	08/02/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	08/02/2023	08/02/2023
					CT2406	Implantable	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	08/02/2023	08/02/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	08/02/2023	08/02/2023
					CT2407	Active implantable	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	08/02/2023	08/02/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	08/02/2023	08/02/2023
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	08/02/2023	08/02/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	08/02/2023	08/02/2023
					CT2727	Active implant devices/ancillaries	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	08/02/2023	08/02/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	08/02/2023	08/02/2023
					CT2750	Implantable cardiac electrical stimulators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	08/02/2023	08/02/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	08/02/2023	08/02/2023
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	08/02/2023	08/02/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	08/02/2023	08/02/2023
					CT2985	Active	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	08/02/2023	08/02/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	08/02/2023	08/02/2023
					CT3001	Implantable metal	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	08/02/2023	08/02/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	08/02/2023	08/02/2023
					CT3002	Implantable synthetic polymer	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	08/02/2023	08/02/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	08/02/2023	08/02/2023
					CT3006	Implantable inorganic material	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	08/02/2023	08/02/2023

							F1901	Additional Surgery	A24	No Apparent Device Problem	08/02/2023	08/02/2023
					CT314	Electrically powered	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	08/02/2023	08/02/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	08/02/2023	08/02/2023
					CT321	Long-term surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	08/02/2023	08/02/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	08/02/2023	08/02/2023
					CT449	Generators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	08/02/2023	08/02/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	08/02/2023	08/02/2023
					CT981	Single use	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	08/02/2023	08/02/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	08/02/2023	08/02/2023
					CT983	Surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	08/02/2023	08/02/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	08/02/2023	08/02/2023
12410	2023		47265	Dual-chamber implantable pacemaker, rate-responsive	CT141	Electrodes/Leads and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	19/02/2023	19/02/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	19/02/2023	19/02/2023
					CT145	Cardiovascular devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	19/02/2023	19/02/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	19/02/2023	19/02/2023
					CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	19/02/2023	19/02/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	19/02/2023	19/02/2023
					CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	19/02/2023	19/02/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	19/02/2023	19/02/2023
					CT233	Surgical	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	19/02/2023	19/02/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	19/02/2023	19/02/2023
					CT2406	Implantable	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	19/02/2023	19/02/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	19/02/2023	19/02/2023
					CT2407	Active implantable	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	19/02/2023	19/02/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	19/02/2023	19/02/2023
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	19/02/2023	19/02/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	19/02/2023	19/02/2023
					CT2727	Active implant devices/ancillaries	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	19/02/2023	19/02/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	19/02/2023	19/02/2023
					CT2750	Implantable cardiac electrical stimulators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	19/02/2023	19/02/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	19/02/2023	19/02/2023
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	19/02/2023	19/02/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	19/02/2023	19/02/2023
					CT2985	Active	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	19/02/2023	19/02/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	19/02/2023	19/02/2023
					CT3001	Implantable metal	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	19/02/2023	19/02/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	19/02/2023	19/02/2023
					CT3002	Implantable synthetic polymer	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	19/02/2023	19/02/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	19/02/2023	19/02/2023
					CT3006	Implantable inorganic material	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	19/02/2023	19/02/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	19/02/2023	19/02/2023
					CT314	Electrically powered	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	19/02/2023	19/02/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	19/02/2023	19/02/2023
					CT321	Long-term surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	19/02/2023	19/02/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	19/02/2023	19/02/2023
					CT449	Generators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	19/02/2023	19/02/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	19/02/2023	19/02/2023
					CT981	Single use	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	19/02/2023	19/02/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	19/02/2023	19/02/2023
					CT983	Surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	19/02/2023	19/02/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	19/02/2023	19/02/2023
12411	2023	All other reportable incidents	35223	Endocardial/interventricular septal pacing lead	CT1684	Cardiac pacemakers and associated devices	E2311	Discomfort	A070909	Oversensing		
							F12	Serious Injury/ Illness/ Impairment	A070909	Oversensing		
12412	2023	All other reportable incidents	60789	Intracardiac pacemaker	CT130	Catheters and associated devices	E0601	Arrhythmia	A0712	Pacing Problem		
							E060106	Heart Block	A0712	Pacing Problem		
							E0618	Non Specific EKG/ECG Changes	A0712	Pacing Problem		
							E1020	Nausea	A0712	Pacing Problem		
							F12	Serious Injury/ Illness/ Impairment	A0712	Pacing Problem		
							F23	Unexpected Medical Intervention	A0712	Pacing Problem		
					CT145	Cardiovascular devices	E0601	Arrhythmia	A0712	Pacing Problem		
							E060106	Heart Block	A0712	Pacing Problem		
							E0618	Non Specific EKG/ECG Changes	A0712	Pacing Problem		
							E1020	Nausea	A0712	Pacing Problem		
							F12	Serious Injury/ Illness/ Impairment	A0712	Pacing Problem		
							F23	Unexpected Medical Intervention	A0712	Pacing Problem		
					CT1684	Cardiac pacemakers and associated devices	E0601	Arrhythmia	A0712	Pacing Problem		
							E060106	Heart Block	A0712	Pacing Problem		
							E0618	Non Specific EKG/ECG Changes	A0712	Pacing Problem		
							E1020	Nausea	A0712	Pacing Problem		
							F12	Serious Injury/ Illness/ Impairment	A0712	Pacing Problem		
							F23	Unexpected Medical Intervention	A0712	Pacing Problem		
					CT1685	Cardiac pacemakers	E0601	Arrhythmia	A0712	Pacing Problem		
							E060106	Heart Block	A0712	Pacing Problem		
							E0618	Non Specific EKG/ECG Changes	A0712	Pacing Problem		
							E1020	Nausea	A0712	Pacing Problem		
							F12	Serious Injury/ Illness/ Impairment	A0712	Pacing Problem		
							F23	Unexpected Medical Intervention	A0712	Pacing Problem		
					CT233	Surgical	E0601	Arrhythmia	A0712	Pacing Problem		
							E060106	Heart Block	A0712	Pacing Problem		
							E0618	Non Specific EKG/ECG Changes	A0712	Pacing Problem		
							E1020	Nausea	A0712	Pacing Problem		
							F12	Serious Injury/ Illness/ Impairment	A0712	Pacing Problem		
							F23	Unexpected Medical Intervention	A0712	Pacing Problem		
					CT2406	Implantable	E0601	Arrhythmia	A0712	Pacing Problem		
							E060106	Heart Block	A0712	Pacing Problem		
							E0618	Non Specific EKG/ECG Changes	A0712	Pacing Problem		
							E1020	Nausea	A0712	Pacing Problem		
							F12	Serious Injury/ Illness/ Impairment	A0712	Pacing Problem		

							F23	Unexpected Medical Intervention	A0712	Pacing Problem		
					CT2407	Active implantable	E0601	Arrhythmia	A0712	Pacing Problem		
							E060106	Heart Block	A0712	Pacing Problem		
							E0618	Non Specific EKG/ECG Changes	A0712	Pacing Problem		
							E1020	Nausea	A0712	Pacing Problem		
							F12	Serious Injury/ Illness/ Impairment	A0712	Pacing Problem		
							F23	Unexpected Medical Intervention	A0712	Pacing Problem		
					CT257	Implantable cardiac pacemakers	E0601	Arrhythmia	A0712	Pacing Problem		
							E060106	Heart Block	A0712	Pacing Problem		
							E0618	Non Specific EKG/ECG Changes	A0712	Pacing Problem		
							E1020	Nausea	A0712	Pacing Problem		
							F12	Serious Injury/ Illness/ Impairment	A0712	Pacing Problem		
							F23	Unexpected Medical Intervention	A0712	Pacing Problem		
					CT2727	Active implant devices/ancillaries	E0601	Arrhythmia	A0712	Pacing Problem		
							E060106	Heart Block	A0712	Pacing Problem		
							E0618	Non Specific EKG/ECG Changes	A0712	Pacing Problem		
							E1020	Nausea	A0712	Pacing Problem		
							F12	Serious Injury/ Illness/ Impairment	A0712	Pacing Problem		
							F23	Unexpected Medical Intervention	A0712	Pacing Problem		
					CT2750	Implantable cardiac electrical stimulators and associated devices	E0601	Arrhythmia	A0712	Pacing Problem		
							E060106	Heart Block	A0712	Pacing Problem		
							E0618	Non Specific EKG/ECG Changes	A0712	Pacing Problem		
							E1020	Nausea	A0712	Pacing Problem		
							F12	Serious Injury/ Illness/ Impairment	A0712	Pacing Problem		
							F23	Unexpected Medical Intervention	A0712	Pacing Problem		
					CT2818	Catheterization/Cannulation/Fluid path devices	E0601	Arrhythmia	A0712	Pacing Problem		
							E060106	Heart Block	A0712	Pacing Problem		
							E0618	Non Specific EKG/ECG Changes	A0712	Pacing Problem		
							E1020	Nausea	A0712	Pacing Problem		
							F12	Serious Injury/ Illness/ Impairment	A0712	Pacing Problem		
							F23	Unexpected Medical Intervention	A0712	Pacing Problem		
					CT2866	Implantable cardiac electrical stimulators	E0601	Arrhythmia	A0712	Pacing Problem		
							E060106	Heart Block	A0712	Pacing Problem		
							E0618	Non Specific EKG/ECG Changes	A0712	Pacing Problem		
							E1020	Nausea	A0712	Pacing Problem		
							F12	Serious Injury/ Illness/ Impairment	A0712	Pacing Problem		
							F23	Unexpected Medical Intervention	A0712	Pacing Problem		
					CT2985	Active	E0601	Arrhythmia	A0712	Pacing Problem		
							E060106	Heart Block	A0712	Pacing Problem		
							E0618	Non Specific EKG/ECG Changes	A0712	Pacing Problem		
							E1020	Nausea	A0712	Pacing Problem		
							F12	Serious Injury/ Illness/ Impairment	A0712	Pacing Problem		
							F23	Unexpected Medical Intervention	A0712	Pacing Problem		
					CT3001	Implantable metal	E0601	Arrhythmia	A0712	Pacing Problem		
							E060106	Heart Block	A0712	Pacing Problem		
							E0618	Non Specific EKG/ECG Changes	A0712	Pacing Problem		
							E1020	Nausea	A0712	Pacing Problem		
							F12	Serious Injury/ Illness/ Impairment	A0712	Pacing Problem		
							F23	Unexpected Medical Intervention	A0712	Pacing Problem		
					CT3002	Implantable synthetic polymer	E0601	Arrhythmia	A0712	Pacing Problem		
							E060106	Heart Block	A0712	Pacing Problem		
							E0618	Non Specific EKG/ECG Changes	A0712	Pacing Problem		
							E1020	Nausea	A0712	Pacing Problem		
							F12	Serious Injury/ Illness/ Impairment	A0712	Pacing Problem		
							F23	Unexpected Medical Intervention	A0712	Pacing Problem		
					CT3006	Implantable inorganic material	E0601	Arrhythmia	A0712	Pacing Problem		
							E060106	Heart Block	A0712	Pacing Problem		
							E0618	Non Specific EKG/ECG Changes	A0712	Pacing Problem		
							E1020	Nausea	A0712	Pacing Problem		
							F12	Serious Injury/ Illness/ Impairment	A0712	Pacing Problem		
							F23	Unexpected Medical Intervention	A0712	Pacing Problem		
					CT314	Electrically powered	E0601	Arrhythmia	A0712	Pacing Problem		
							E060106	Heart Block	A0712	Pacing Problem		
							E0618	Non Specific EKG/ECG Changes	A0712	Pacing Problem		
							E1020	Nausea	A0712	Pacing Problem		
							F12	Serious Injury/ Illness/ Impairment	A0712	Pacing Problem		
							F23	Unexpected Medical Intervention	A0712	Pacing Problem		
					CT321	Long-term surgical invasive	E0601	Arrhythmia	A0712	Pacing Problem		
							E060106	Heart Block	A0712	Pacing Problem		
							E0618	Non Specific EKG/ECG Changes	A0712	Pacing Problem		
							E1020	Nausea	A0712	Pacing Problem		
							F12	Serious Injury/ Illness/ Impairment	A0712	Pacing Problem		
							F23	Unexpected Medical Intervention	A0712	Pacing Problem		
					CT981	Single use	E0601	Arrhythmia	A0712	Pacing Problem		
							E060106	Heart Block	A0712	Pacing Problem		
							E0618	Non Specific EKG/ECG Changes	A0712	Pacing Problem		
							E1020	Nausea	A0712	Pacing Problem		
							F12	Serious Injury/ Illness/ Impairment	A0712	Pacing Problem		
							F23	Unexpected Medical Intervention	A0712	Pacing Problem		
					CT983	Surgical invasive	E0601	Arrhythmia	A0712	Pacing Problem		
							E060106	Heart Block	A0712	Pacing Problem		
							E0618	Non Specific EKG/ECG Changes	A0712	Pacing Problem		
							E1020	Nausea	A0712	Pacing Problem		
							F12	Serious Injury/ Illness/ Impairment	A0712	Pacing Problem		
							F23	Unexpected Medical Intervention	A0712	Pacing Problem		
12413	2023	All other reportable incidents	35223	Endocardial/interventricular septal pacing lead	CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A072202	Low Impedance		
							F26	No Health Consequences or Impact	A072202	Low Impedance		
12414	2023	All other reportable incidents	35223	Endocardial/interventricular septal pacing lead	CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A07	Electrical /Electronic Property Problem		

									A0712	Pacing Problem		
									A072201	High Impedance		
									A090801	Signal Artifact/Noise		
							F26	No Health Consequences or Impact	A07	Electrical /Electronic Property Problem		
									A0712	Pacing Problem		
									A072201	High Impedance		
									A090801	Signal Artifact/Noise		
12415	2023	All other reportable incidents	35223	Endocardial/interventricular septal pacing lead	CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070909	Oversensing		
									A090801	Signal Artifact/Noise		
							F26	No Health Consequences or Impact	A070909	Oversensing		
									A090801	Signal Artifact/Noise		
12416	2023	All other reportable incidents	47265	Dual-chamber implantable pacemaker, rate-responsive	CT141	Electrodes/Leads and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
									A071102	Failure to Interrogate		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
									A071102	Failure to Interrogate		
					CT145	Cardiovascular devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
									A071102	Failure to Interrogate		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
									A071102	Failure to Interrogate		
					CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
									A071102	Failure to Interrogate		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
									A071102	Failure to Interrogate		
					CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
									A071102	Failure to Interrogate		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
									A071102	Failure to Interrogate		
					CT233	Surgical	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
									A071102	Failure to Interrogate		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
									A071102	Failure to Interrogate		
					CT2406	Implantable	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
									A071102	Failure to Interrogate		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
									A071102	Failure to Interrogate		
					CT2407	Active implantable	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
									A071102	Failure to Interrogate		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
									A071102	Failure to Interrogate		
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
									A071102	Failure to Interrogate		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
									A071102	Failure to Interrogate		
					CT2727	Active implant devices/ancillaries	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
									A071102	Failure to Interrogate		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
									A071102	Failure to Interrogate		
					CT2750	Implantable cardiac electrical stimulators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
									A071102	Failure to Interrogate		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
									A071102	Failure to Interrogate		
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
									A071102	Failure to Interrogate		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
									A071102	Failure to Interrogate		
					CT2985	Active	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
									A071102	Failure to Interrogate		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
									A071102	Failure to Interrogate		
					CT3001	Implantable metal	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
									A071102	Failure to Interrogate		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
									A071102	Failure to Interrogate		
					CT3002	Implantable synthetic polymer	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
									A071102	Failure to Interrogate		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
									A071102	Failure to Interrogate		
					CT3006	Implantable inorganic material	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
									A071102	Failure to Interrogate		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
									A071102	Failure to Interrogate		
					CT314	Electrically powered	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
									A071102	Failure to Interrogate		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
									A071102	Failure to Interrogate		
					CT321	Long-term surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
									A071102	Failure to Interrogate		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
									A071102	Failure to Interrogate		
					CT449	Generators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
									A071102	Failure to Interrogate		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
									A071102	Failure to Interrogate		
					CT981	Single use	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
									A071102	Failure to Interrogate		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
									A071102	Failure to Interrogate		
					CT983	Surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		

								A071102	Failure to Interrogate			
						F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery			
								A071102	Failure to Interrogate			
12417	2023	All other reportable incidents	47265	Dual-chamber implantable pacemaker, rate-responsive	CT141	Electrodes/Leads and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	09/11/2022	09/11/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	09/11/2022	09/11/2022
					CT145	Cardiovascular devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	09/11/2022	09/11/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	09/11/2022	09/11/2022
					CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	09/11/2022	09/11/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	09/11/2022	09/11/2022
					CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	09/11/2022	09/11/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	09/11/2022	09/11/2022
					CT233	Surgical	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	09/11/2022	09/11/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	09/11/2022	09/11/2022
					CT2406	Implantable	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	09/11/2022	09/11/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	09/11/2022	09/11/2022
					CT2407	Active implantable	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	09/11/2022	09/11/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	09/11/2022	09/11/2022
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	09/11/2022	09/11/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	09/11/2022	09/11/2022
					CT2727	Active implant devices/ancillaries	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	09/11/2022	09/11/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	09/11/2022	09/11/2022
					CT2750	Implantable cardiac electrical stimulators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	09/11/2022	09/11/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	09/11/2022	09/11/2022
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	09/11/2022	09/11/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	09/11/2022	09/11/2022
					CT2985	Active	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	09/11/2022	09/11/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	09/11/2022	09/11/2022
					CT3001	Implantable metal	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	09/11/2022	09/11/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	09/11/2022	09/11/2022
					CT3002	Implantable synthetic polymer	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	09/11/2022	09/11/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	09/11/2022	09/11/2022
					CT3006	Implantable inorganic material	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	09/11/2022	09/11/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	09/11/2022	09/11/2022
					CT314	Electrically powered	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	09/11/2022	09/11/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	09/11/2022	09/11/2022
					CT321	Long-term surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	09/11/2022	09/11/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	09/11/2022	09/11/2022
					CT449	Generators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	09/11/2022	09/11/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	09/11/2022	09/11/2022
					CT981	Single use	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	09/11/2022	09/11/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	09/11/2022	09/11/2022
					CT983	Surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	09/11/2022	09/11/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	09/11/2022	09/11/2022
12418	2023	All other reportable incidents	47265	Dual-chamber implantable pacemaker, rate-responsive	CT141	Electrodes/Leads and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT145	Cardiovascular devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT233	Surgical	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT2406	Implantable	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT2407	Active implantable	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT2727	Active implant devices/ancillaries	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT2750	Implantable cardiac electrical stimulators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT2985	Active	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT3001	Implantable metal	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT3002	Implantable synthetic polymer	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT3006	Implantable inorganic material	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT314	Electrically powered	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT321	Long-term surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT449	Generators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT981	Single use	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT983	Surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
12419	2023	All other reportable incidents	35852	Single-chamber implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
12420	2023	All other reportable incidents	35223	Endocardial/interventricular septal pacing lead	CT1684	Cardiac pacemakers and associated devices	E0112	Dizziness	A0701	Capturing Problem	02/03/2023	02/03/2023

								A070909	Oversensing	02/03/2023	02/03/2023
								A072202	Low Impedance	02/03/2023	02/03/2023
						F1901	Additional Surgery	A0701	Capturing Problem	02/03/2023	02/03/2023
								A070909	Oversensing	02/03/2023	02/03/2023
								A072202	Low Impedance	02/03/2023	02/03/2023
12421	2023	All other reportable incidents	47263	Cardiac resynchronization therapy implantable pacemaker	CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	14/12/2022	14/12/2022
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	14/12/2022	14/12/2022
							F12	Serious Injury/ Illness/ Impairment	A070504	Premature Discharge of Battery	14/12/2022	14/12/2022
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	14/12/2022	14/12/2022
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	14/12/2022	14/12/2022
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	14/12/2022	14/12/2022
							F12	Serious Injury/ Illness/ Impairment	A070504	Premature Discharge of Battery	14/12/2022	14/12/2022
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	14/12/2022	14/12/2022
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	14/12/2022	14/12/2022
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	14/12/2022	14/12/2022
							F12	Serious Injury/ Illness/ Impairment	A070504	Premature Discharge of Battery	14/12/2022	14/12/2022
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	14/12/2022	14/12/2022
12422	2023		47265	Dual-chamber implantable pacemaker, rate-responsive	CT141	Electrodes/Leads and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	25/02/2023	25/02/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	25/02/2023	25/02/2023
					CT145	Cardiovascular devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	25/02/2023	25/02/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	25/02/2023	25/02/2023
					CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	25/02/2023	25/02/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	25/02/2023	25/02/2023
					CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	25/02/2023	25/02/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	25/02/2023	25/02/2023
					CT233	Surgical	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	25/02/2023	25/02/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	25/02/2023	25/02/2023
					CT2406	Implantable	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	25/02/2023	25/02/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	25/02/2023	25/02/2023
					CT2407	Active implantable	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	25/02/2023	25/02/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	25/02/2023	25/02/2023
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	25/02/2023	25/02/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	25/02/2023	25/02/2023
					CT2727	Active implant devices/ancillaries	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	25/02/2023	25/02/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	25/02/2023	25/02/2023
					CT2750	Implantable cardiac electrical stimulators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	25/02/2023	25/02/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	25/02/2023	25/02/2023
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	25/02/2023	25/02/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	25/02/2023	25/02/2023
					CT2985	Active	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	25/02/2023	25/02/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	25/02/2023	25/02/2023
					CT3001	Implantable metal	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	25/02/2023	25/02/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	25/02/2023	25/02/2023
					CT3002	Implantable synthetic polymer	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	25/02/2023	25/02/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	25/02/2023	25/02/2023
					CT3006	Implantable inorganic material	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	25/02/2023	25/02/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	25/02/2023	25/02/2023
					CT314	Electrically powered	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	25/02/2023	25/02/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	25/02/2023	25/02/2023
					CT321	Long-term surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	25/02/2023	25/02/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	25/02/2023	25/02/2023
					CT449	Generators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	25/02/2023	25/02/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	25/02/2023	25/02/2023
					CT981	Single use	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	25/02/2023	25/02/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	25/02/2023	25/02/2023
					CT983	Surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	25/02/2023	25/02/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	25/02/2023	25/02/2023
12423	2023		47265	Dual-chamber implantable pacemaker, rate-responsive	CT141	Electrodes/Leads and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	25/02/2023	25/02/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	25/02/2023	25/02/2023
					CT145	Cardiovascular devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	25/02/2023	25/02/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	25/02/2023	25/02/2023
					CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	25/02/2023	25/02/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	25/02/2023	25/02/2023
					CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	25/02/2023	25/02/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	25/02/2023	25/02/2023
					CT233	Surgical	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	25/02/2023	25/02/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	25/02/2023	25/02/2023
					CT2406	Implantable	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	25/02/2023	25/02/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	25/02/2023	25/02/2023
					CT2407	Active implantable	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	25/02/2023	25/02/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	25/02/2023	25/02/2023
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	25/02/2023	25/02/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	25/02/2023	25/02/2023
					CT2727	Active implant devices/ancillaries	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	25/02/2023	25/02/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	25/02/2023	25/02/2023
					CT2750	Implantable cardiac electrical stimulators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	25/02/2023	25/02/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	25/02/2023	25/02/2023
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	25/02/2023	25/02/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	25/02/2023	25/02/2023
					CT2985	Active	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	25/02/2023	25/02/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	25/02/2023	25/02/2023
					CT3001	Implantable metal	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	25/02/2023	25/02/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	25/02/2023	25/02/2023
					CT3002	Implantable synthetic polymer	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	25/02/2023	25/02/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	25/02/2023	25/02/2023
					CT3006	Implantable inorganic material	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	25/02/2023	25/02/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	25/02/2023	25/02/2023
					CT314	Electrically powered	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	25/02/2023	25/02/2023

							F1901	Additional Surgery	A24	No Apparent Device Problem	25/02/2023	25/02/2023
					CT321	Long-term surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	25/02/2023	25/02/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	25/02/2023	25/02/2023
					CT449	Generators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	25/02/2023	25/02/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	25/02/2023	25/02/2023
					CT981	Single use	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	25/02/2023	25/02/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	25/02/2023	25/02/2023
					CT983	Surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	25/02/2023	25/02/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	25/02/2023	25/02/2023
12424	2023	All other reportable incidents	35852	Single-chamber implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
									A0908	Incorrect, Inadequate or Imprecise Result or Readings		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
									A0908	Incorrect, Inadequate or Imprecise Result or Readings		
12425	2023	All other reportable incidents	47265	Dual-chamber implantable pacemaker, rate-responsive	CT141	Electrodes/Leads and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem	21/02/2023	21/02/2023
									A070504	Premature Discharge of Battery	21/02/2023	21/02/2023
							F08	Hospitalization or Prolonged Hospitalization	A0705	Battery Problem	21/02/2023	21/02/2023
									A070504	Premature Discharge of Battery	21/02/2023	21/02/2023
							F1905	Device Revision or Replacement	A0705	Battery Problem	21/02/2023	21/02/2023
									A070504	Premature Discharge of Battery	21/02/2023	21/02/2023
					CT145	Cardiovascular devices	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem	21/02/2023	21/02/2023
									A070504	Premature Discharge of Battery	21/02/2023	21/02/2023
							F08	Hospitalization or Prolonged Hospitalization	A0705	Battery Problem	21/02/2023	21/02/2023
									A070504	Premature Discharge of Battery	21/02/2023	21/02/2023
							F1905	Device Revision or Replacement	A0705	Battery Problem	21/02/2023	21/02/2023
									A070504	Premature Discharge of Battery	21/02/2023	21/02/2023
					CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem	21/02/2023	21/02/2023
									A070504	Premature Discharge of Battery	21/02/2023	21/02/2023
							F08	Hospitalization or Prolonged Hospitalization	A0705	Battery Problem	21/02/2023	21/02/2023
									A070504	Premature Discharge of Battery	21/02/2023	21/02/2023
							F1905	Device Revision or Replacement	A0705	Battery Problem	21/02/2023	21/02/2023
									A070504	Premature Discharge of Battery	21/02/2023	21/02/2023
					CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem	21/02/2023	21/02/2023
									A070504	Premature Discharge of Battery	21/02/2023	21/02/2023
							F08	Hospitalization or Prolonged Hospitalization	A0705	Battery Problem	21/02/2023	21/02/2023
									A070504	Premature Discharge of Battery	21/02/2023	21/02/2023
							F1905	Device Revision or Replacement	A0705	Battery Problem	21/02/2023	21/02/2023
									A070504	Premature Discharge of Battery	21/02/2023	21/02/2023
					CT233	Surgical	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem	21/02/2023	21/02/2023
									A070504	Premature Discharge of Battery	21/02/2023	21/02/2023
							F08	Hospitalization or Prolonged Hospitalization	A0705	Battery Problem	21/02/2023	21/02/2023
									A070504	Premature Discharge of Battery	21/02/2023	21/02/2023
							F1905	Device Revision or Replacement	A0705	Battery Problem	21/02/2023	21/02/2023
									A070504	Premature Discharge of Battery	21/02/2023	21/02/2023
					CT2406	Implantable	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem	21/02/2023	21/02/2023
									A070504	Premature Discharge of Battery	21/02/2023	21/02/2023
							F08	Hospitalization or Prolonged Hospitalization	A0705	Battery Problem	21/02/2023	21/02/2023
									A070504	Premature Discharge of Battery	21/02/2023	21/02/2023
							F1905	Device Revision or Replacement	A0705	Battery Problem	21/02/2023	21/02/2023
									A070504	Premature Discharge of Battery	21/02/2023	21/02/2023
					CT2407	Active implantable	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem	21/02/2023	21/02/2023
									A070504	Premature Discharge of Battery	21/02/2023	21/02/2023
							F08	Hospitalization or Prolonged Hospitalization	A0705	Battery Problem	21/02/2023	21/02/2023
									A070504	Premature Discharge of Battery	21/02/2023	21/02/2023
							F1905	Device Revision or Replacement	A0705	Battery Problem	21/02/2023	21/02/2023
									A070504	Premature Discharge of Battery	21/02/2023	21/02/2023
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem	21/02/2023	21/02/2023
									A070504	Premature Discharge of Battery	21/02/2023	21/02/2023
							F08	Hospitalization or Prolonged Hospitalization	A0705	Battery Problem	21/02/2023	21/02/2023
									A070504	Premature Discharge of Battery	21/02/2023	21/02/2023
							F1905	Device Revision or Replacement	A0705	Battery Problem	21/02/2023	21/02/2023
									A070504	Premature Discharge of Battery	21/02/2023	21/02/2023
					CT2727	Active implant devices/ancillaries	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem	21/02/2023	21/02/2023
									A070504	Premature Discharge of Battery	21/02/2023	21/02/2023
							F08	Hospitalization or Prolonged Hospitalization	A0705	Battery Problem	21/02/2023	21/02/2023
									A070504	Premature Discharge of Battery	21/02/2023	21/02/2023
							F1905	Device Revision or Replacement	A0705	Battery Problem	21/02/2023	21/02/2023
									A070504	Premature Discharge of Battery	21/02/2023	21/02/2023
					CT2750	Implantable cardiac electrical stimulators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem	21/02/2023	21/02/2023
									A070504	Premature Discharge of Battery	21/02/2023	21/02/2023
							F08	Hospitalization or Prolonged Hospitalization	A0705	Battery Problem	21/02/2023	21/02/2023
									A070504	Premature Discharge of Battery	21/02/2023	21/02/2023
							F1905	Device Revision or Replacement	A0705	Battery Problem	21/02/2023	21/02/2023
									A070504	Premature Discharge of Battery	21/02/2023	21/02/2023
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem	21/02/2023	21/02/2023
									A070504	Premature Discharge of Battery	21/02/2023	21/02/2023
							F08	Hospitalization or Prolonged Hospitalization	A0705	Battery Problem	21/02/2023	21/02/2023
									A070504	Premature Discharge of Battery	21/02/2023	21/02/2023
							F1905	Device Revision or Replacement	A0705	Battery Problem	21/02/2023	21/02/2023
									A070504	Premature Discharge of Battery	21/02/2023	21/02/2023
					CT2985	Active	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem	21/02/2023	21/02/2023
									A070504	Premature Discharge of Battery	21/02/2023	21/02/2023
							F08	Hospitalization or Prolonged Hospitalization	A0705	Battery Problem	21/02/2023	21/02/2023
									A070504	Premature Discharge of Battery	21/02/2023	21/02/2023
							F1905	Device Revision or Replacement	A0705	Battery Problem	21/02/2023	21/02/2023
									A070504	Premature Discharge of Battery	21/02/2023	21/02/2023
					CT3001	Implantable metal	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem	21/02/2023	21/02/2023

									A070504	Premature Discharge of Battery	21/02/2023	21/02/2023
							F08	Hospitalization or Prolonged Hospitalization	A0705	Battery Problem	21/02/2023	21/02/2023
									A070504	Premature Discharge of Battery	21/02/2023	21/02/2023
							F1905	Device Revision or Replacement	A0705	Battery Problem	21/02/2023	21/02/2023
									A070504	Premature Discharge of Battery	21/02/2023	21/02/2023
					CT3002	Implantable synthetic polymer	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem	21/02/2023	21/02/2023
									A070504	Premature Discharge of Battery	21/02/2023	21/02/2023
							F08	Hospitalization or Prolonged Hospitalization	A0705	Battery Problem	21/02/2023	21/02/2023
									A070504	Premature Discharge of Battery	21/02/2023	21/02/2023
							F1905	Device Revision or Replacement	A0705	Battery Problem	21/02/2023	21/02/2023
									A070504	Premature Discharge of Battery	21/02/2023	21/02/2023
					CT3006	Implantable inorganic material	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem	21/02/2023	21/02/2023
									A070504	Premature Discharge of Battery	21/02/2023	21/02/2023
							F08	Hospitalization or Prolonged Hospitalization	A0705	Battery Problem	21/02/2023	21/02/2023
									A070504	Premature Discharge of Battery	21/02/2023	21/02/2023
							F1905	Device Revision or Replacement	A0705	Battery Problem	21/02/2023	21/02/2023
									A070504	Premature Discharge of Battery	21/02/2023	21/02/2023
					CT314	Electrically powered	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem	21/02/2023	21/02/2023
									A070504	Premature Discharge of Battery	21/02/2023	21/02/2023
							F08	Hospitalization or Prolonged Hospitalization	A0705	Battery Problem	21/02/2023	21/02/2023
									A070504	Premature Discharge of Battery	21/02/2023	21/02/2023
							F1905	Device Revision or Replacement	A0705	Battery Problem	21/02/2023	21/02/2023
									A070504	Premature Discharge of Battery	21/02/2023	21/02/2023
					CT321	Long-term surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem	21/02/2023	21/02/2023
									A070504	Premature Discharge of Battery	21/02/2023	21/02/2023
							F08	Hospitalization or Prolonged Hospitalization	A0705	Battery Problem	21/02/2023	21/02/2023
									A070504	Premature Discharge of Battery	21/02/2023	21/02/2023
							F1905	Device Revision or Replacement	A0705	Battery Problem	21/02/2023	21/02/2023
									A070504	Premature Discharge of Battery	21/02/2023	21/02/2023
					CT449	Generators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem	21/02/2023	21/02/2023
									A070504	Premature Discharge of Battery	21/02/2023	21/02/2023
							F08	Hospitalization or Prolonged Hospitalization	A0705	Battery Problem	21/02/2023	21/02/2023
									A070504	Premature Discharge of Battery	21/02/2023	21/02/2023
							F1905	Device Revision or Replacement	A0705	Battery Problem	21/02/2023	21/02/2023
									A070504	Premature Discharge of Battery	21/02/2023	21/02/2023
					CT981	Single use	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem	21/02/2023	21/02/2023
									A070504	Premature Discharge of Battery	21/02/2023	21/02/2023
							F08	Hospitalization or Prolonged Hospitalization	A0705	Battery Problem	21/02/2023	21/02/2023
									A070504	Premature Discharge of Battery	21/02/2023	21/02/2023
							F1905	Device Revision or Replacement	A0705	Battery Problem	21/02/2023	21/02/2023
									A070504	Premature Discharge of Battery	21/02/2023	21/02/2023
					CT983	Surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem	21/02/2023	21/02/2023
									A070504	Premature Discharge of Battery	21/02/2023	21/02/2023
							F08	Hospitalization or Prolonged Hospitalization	A0705	Battery Problem	21/02/2023	21/02/2023
									A070504	Premature Discharge of Battery	21/02/2023	21/02/2023
							F1905	Device Revision or Replacement	A0705	Battery Problem	21/02/2023	21/02/2023
									A070504	Premature Discharge of Battery	21/02/2023	21/02/2023
12426	2023		47265	Dual-chamber implantable pacemaker, rate-responsive	CT141	Electrodes/Leads and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	29/01/2023	29/01/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	29/01/2023	29/01/2023
					CT145	Cardiovascular devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	29/01/2023	29/01/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	29/01/2023	29/01/2023
					CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	29/01/2023	29/01/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	29/01/2023	29/01/2023
					CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	29/01/2023	29/01/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	29/01/2023	29/01/2023
					CT233	Surgical	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	29/01/2023	29/01/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	29/01/2023	29/01/2023
					CT2406	Implantable	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	29/01/2023	29/01/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	29/01/2023	29/01/2023
					CT2407	Active implantable	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	29/01/2023	29/01/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	29/01/2023	29/01/2023
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	29/01/2023	29/01/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	29/01/2023	29/01/2023
					CT2727	Active implant devices/ancillaries	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	29/01/2023	29/01/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	29/01/2023	29/01/2023
					CT2750	Implantable cardiac electrical stimulators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	29/01/2023	29/01/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	29/01/2023	29/01/2023
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	29/01/2023	29/01/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	29/01/2023	29/01/2023
					CT2985	Active	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	29/01/2023	29/01/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	29/01/2023	29/01/2023
					CT3001	Implantable metal	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	29/01/2023	29/01/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	29/01/2023	29/01/2023
					CT3002	Implantable synthetic polymer	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	29/01/2023	29/01/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	29/01/2023	29/01/2023
					CT3006	Implantable inorganic material	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	29/01/2023	29/01/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	29/01/2023	29/01/2023
					CT314	Electrically powered	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	29/01/2023	29/01/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	29/01/2023	29/01/2023
					CT321	Long-term surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	29/01/2023	29/01/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	29/01/2023	29/01/2023
					CT449	Generators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	29/01/2023	29/01/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	29/01/2023	29/01/2023
					CT981	Single use	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	29/01/2023	29/01/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	29/01/2023	29/01/2023
					CT983	Surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	29/01/2023	29/01/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	29/01/2023	29/01/2023
12427	2023		47265	Dual-chamber implantable pacemaker, rate-	CT141	Electrodes/Leads and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	17/02/2023	17/02/2023

				responsive			F1901	Additional Surgery	A24	No Apparent Device Problem	17/02/2023	17/02/2023
					CT145	Cardiovascular devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	17/02/2023	17/02/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	17/02/2023	17/02/2023
					CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	17/02/2023	17/02/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	17/02/2023	17/02/2023
					CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	17/02/2023	17/02/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	17/02/2023	17/02/2023
					CT233	Surgical	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	17/02/2023	17/02/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	17/02/2023	17/02/2023
					CT2406	Implantable	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	17/02/2023	17/02/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	17/02/2023	17/02/2023
					CT2407	Active implantable	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	17/02/2023	17/02/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	17/02/2023	17/02/2023
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	17/02/2023	17/02/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	17/02/2023	17/02/2023
					CT2727	Active implant devices/ancillaries	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	17/02/2023	17/02/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	17/02/2023	17/02/2023
					CT2750	Implantable cardiac electrical stimulators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	17/02/2023	17/02/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	17/02/2023	17/02/2023
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	17/02/2023	17/02/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	17/02/2023	17/02/2023
					CT2985	Active	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	17/02/2023	17/02/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	17/02/2023	17/02/2023
					CT3001	Implantable metal	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	17/02/2023	17/02/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	17/02/2023	17/02/2023
					CT3002	Implantable synthetic polymer	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	17/02/2023	17/02/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	17/02/2023	17/02/2023
					CT3006	Implantable inorganic material	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	17/02/2023	17/02/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	17/02/2023	17/02/2023
					CT314	Electrically powered	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	17/02/2023	17/02/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	17/02/2023	17/02/2023
					CT321	Long-term surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	17/02/2023	17/02/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	17/02/2023	17/02/2023
					CT449	Generators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	17/02/2023	17/02/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	17/02/2023	17/02/2023
					CT981	Single use	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	17/02/2023	17/02/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	17/02/2023	17/02/2023
					CT983	Surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	17/02/2023	17/02/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	17/02/2023	17/02/2023
12428	2023		47265	Dual-chamber implantable pacemaker, rate-responsive	CT141	Electrodes/Leads and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	09/03/2023	09/03/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	09/03/2023	09/03/2023
					CT145	Cardiovascular devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	09/03/2023	09/03/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	09/03/2023	09/03/2023
					CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	09/03/2023	09/03/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	09/03/2023	09/03/2023
					CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	09/03/2023	09/03/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	09/03/2023	09/03/2023
					CT233	Surgical	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	09/03/2023	09/03/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	09/03/2023	09/03/2023
					CT2406	Implantable	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	09/03/2023	09/03/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	09/03/2023	09/03/2023
					CT2407	Active implantable	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	09/03/2023	09/03/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	09/03/2023	09/03/2023
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	09/03/2023	09/03/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	09/03/2023	09/03/2023
					CT2727	Active implant devices/ancillaries	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	09/03/2023	09/03/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	09/03/2023	09/03/2023
					CT2750	Implantable cardiac electrical stimulators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	09/03/2023	09/03/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	09/03/2023	09/03/2023
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	09/03/2023	09/03/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	09/03/2023	09/03/2023
					CT2985	Active	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	09/03/2023	09/03/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	09/03/2023	09/03/2023
					CT3001	Implantable metal	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	09/03/2023	09/03/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	09/03/2023	09/03/2023
					CT3002	Implantable synthetic polymer	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	09/03/2023	09/03/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	09/03/2023	09/03/2023
					CT3006	Implantable inorganic material	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	09/03/2023	09/03/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	09/03/2023	09/03/2023
					CT314	Electrically powered	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	09/03/2023	09/03/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	09/03/2023	09/03/2023
					CT321	Long-term surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	09/03/2023	09/03/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	09/03/2023	09/03/2023
					CT449	Generators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	09/03/2023	09/03/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	09/03/2023	09/03/2023
					CT981	Single use	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	09/03/2023	09/03/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	09/03/2023	09/03/2023
					CT983	Surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	09/03/2023	09/03/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	09/03/2023	09/03/2023
12429	2023		47265	Dual-chamber implantable pacemaker, rate-responsive	CT141	Electrodes/Leads and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	20/02/2023	20/02/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	20/02/2023	20/02/2023
					CT145	Cardiovascular devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	20/02/2023	20/02/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	20/02/2023	20/02/2023
					CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	20/02/2023	20/02/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	20/02/2023	20/02/2023
					CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	20/02/2023	20/02/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	20/02/2023	20/02/2023
					CT233	Surgical	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	20/02/2023	20/02/2023

							F1901	Additional Surgery	A24	No Apparent Device Problem	20/02/2023	20/02/2023
					CT2406	Implantable	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	20/02/2023	20/02/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	20/02/2023	20/02/2023
					CT2407	Active implantable	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	20/02/2023	20/02/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	20/02/2023	20/02/2023
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	20/02/2023	20/02/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	20/02/2023	20/02/2023
					CT2727	Active implant devices/ancillaries	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	20/02/2023	20/02/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	20/02/2023	20/02/2023
					CT2750	Implantable cardiac electrical stimulators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	20/02/2023	20/02/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	20/02/2023	20/02/2023
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	20/02/2023	20/02/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	20/02/2023	20/02/2023
					CT2985	Active	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	20/02/2023	20/02/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	20/02/2023	20/02/2023
					CT3001	Implantable metal	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	20/02/2023	20/02/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	20/02/2023	20/02/2023
					CT3002	Implantable synthetic polymer	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	20/02/2023	20/02/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	20/02/2023	20/02/2023
					CT3006	Implantable inorganic material	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	20/02/2023	20/02/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	20/02/2023	20/02/2023
					CT314	Electrically powered	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	20/02/2023	20/02/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	20/02/2023	20/02/2023
					CT321	Long-term surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	20/02/2023	20/02/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	20/02/2023	20/02/2023
					CT449	Generators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	20/02/2023	20/02/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	20/02/2023	20/02/2023
					CT981	Single use	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	20/02/2023	20/02/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	20/02/2023	20/02/2023
					CT983	Surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	20/02/2023	20/02/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	20/02/2023	20/02/2023
12430	2023	All other reportable incidents	47265	Dual-chamber implantable pacemaker, rate-responsive	CT141	Electrodes/Leads and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	05/02/2023	05/02/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	05/02/2023	05/02/2023
					CT145	Cardiovascular devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	05/02/2023	05/02/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	05/02/2023	05/02/2023
					CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	05/02/2023	05/02/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	05/02/2023	05/02/2023
					CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	05/02/2023	05/02/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	05/02/2023	05/02/2023
					CT233	Surgical	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	05/02/2023	05/02/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	05/02/2023	05/02/2023
					CT2406	Implantable	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	05/02/2023	05/02/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	05/02/2023	05/02/2023
					CT2407	Active implantable	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	05/02/2023	05/02/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	05/02/2023	05/02/2023
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	05/02/2023	05/02/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	05/02/2023	05/02/2023
					CT2727	Active implant devices/ancillaries	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	05/02/2023	05/02/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	05/02/2023	05/02/2023
					CT2750	Implantable cardiac electrical stimulators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	05/02/2023	05/02/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	05/02/2023	05/02/2023
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	05/02/2023	05/02/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	05/02/2023	05/02/2023
					CT2985	Active	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	05/02/2023	05/02/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	05/02/2023	05/02/2023
					CT3001	Implantable metal	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	05/02/2023	05/02/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	05/02/2023	05/02/2023
					CT3002	Implantable synthetic polymer	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	05/02/2023	05/02/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	05/02/2023	05/02/2023
					CT3006	Implantable inorganic material	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	05/02/2023	05/02/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	05/02/2023	05/02/2023
					CT314	Electrically powered	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	05/02/2023	05/02/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	05/02/2023	05/02/2023
					CT321	Long-term surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	05/02/2023	05/02/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	05/02/2023	05/02/2023
					CT449	Generators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	05/02/2023	05/02/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	05/02/2023	05/02/2023
					CT981	Single use	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	05/02/2023	05/02/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	05/02/2023	05/02/2023
					CT983	Surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	05/02/2023	05/02/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	05/02/2023	05/02/2023
12431	2023	All other reportable incidents	47265	Dual-chamber implantable pacemaker, rate-responsive	CT141	Electrodes/Leads and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	21/02/2023	21/02/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	21/02/2023	21/02/2023
					CT145	Cardiovascular devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	21/02/2023	21/02/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	21/02/2023	21/02/2023
					CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	21/02/2023	21/02/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	21/02/2023	21/02/2023
					CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	21/02/2023	21/02/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	21/02/2023	21/02/2023
					CT233	Surgical	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	21/02/2023	21/02/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	21/02/2023	21/02/2023
					CT2406	Implantable	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	21/02/2023	21/02/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	21/02/2023	21/02/2023
					CT2407	Active implantable	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	21/02/2023	21/02/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	21/02/2023	21/02/2023
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	21/02/2023	21/02/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	21/02/2023	21/02/2023
					CT2727	Active implant devices/ancillaries	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	21/02/2023	21/02/2023

							F1901	Additional Surgery	A24	No Apparent Device Problem	21/02/2023	21/02/2023
					CT2750	Implantable cardiac electrical stimulators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	21/02/2023	21/02/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	21/02/2023	21/02/2023
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	21/02/2023	21/02/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	21/02/2023	21/02/2023
					CT2985	Active	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	21/02/2023	21/02/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	21/02/2023	21/02/2023
					CT3001	Implantable metal	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	21/02/2023	21/02/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	21/02/2023	21/02/2023
					CT3002	Implantable synthetic polymer	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	21/02/2023	21/02/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	21/02/2023	21/02/2023
					CT3006	Implantable inorganic material	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	21/02/2023	21/02/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	21/02/2023	21/02/2023
					CT314	Electrically powered	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	21/02/2023	21/02/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	21/02/2023	21/02/2023
					CT321	Long-term surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	21/02/2023	21/02/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	21/02/2023	21/02/2023
					CT449	Generators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	21/02/2023	21/02/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	21/02/2023	21/02/2023
					CT981	Single use	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	21/02/2023	21/02/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	21/02/2023	21/02/2023
					CT983	Surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	21/02/2023	21/02/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	21/02/2023	21/02/2023
12432	2023		47265	Dual-chamber implantable pacemaker, rate-responsive	CT141	Electrodes/Leads and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	05/03/2023	05/03/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	05/03/2023	05/03/2023
					CT145	Cardiovascular devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	05/03/2023	05/03/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	05/03/2023	05/03/2023
					CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	05/03/2023	05/03/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	05/03/2023	05/03/2023
					CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	05/03/2023	05/03/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	05/03/2023	05/03/2023
					CT233	Surgical	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	05/03/2023	05/03/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	05/03/2023	05/03/2023
					CT2406	Implantable	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	05/03/2023	05/03/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	05/03/2023	05/03/2023
					CT2407	Active implantable	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	05/03/2023	05/03/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	05/03/2023	05/03/2023
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	05/03/2023	05/03/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	05/03/2023	05/03/2023
					CT2727	Active implant devices/ancillaries	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	05/03/2023	05/03/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	05/03/2023	05/03/2023
					CT2750	Implantable cardiac electrical stimulators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	05/03/2023	05/03/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	05/03/2023	05/03/2023
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	05/03/2023	05/03/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	05/03/2023	05/03/2023
					CT2985	Active	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	05/03/2023	05/03/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	05/03/2023	05/03/2023
					CT3001	Implantable metal	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	05/03/2023	05/03/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	05/03/2023	05/03/2023
					CT3002	Implantable synthetic polymer	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	05/03/2023	05/03/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	05/03/2023	05/03/2023
					CT3006	Implantable inorganic material	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	05/03/2023	05/03/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	05/03/2023	05/03/2023
					CT314	Electrically powered	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	05/03/2023	05/03/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	05/03/2023	05/03/2023
					CT321	Long-term surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	05/03/2023	05/03/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	05/03/2023	05/03/2023
					CT449	Generators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	05/03/2023	05/03/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	05/03/2023	05/03/2023
					CT981	Single use	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	05/03/2023	05/03/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	05/03/2023	05/03/2023
					CT983	Surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	05/03/2023	05/03/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	05/03/2023	05/03/2023
12433	2023	All other reportable incidents	47265	Dual-chamber implantable pacemaker, rate-responsive	CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F27	Problem Identified During Non-Clinical Procedure	A070504	Premature Discharge of Battery		
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F27	Problem Identified During Non-Clinical Procedure	A070504	Premature Discharge of Battery		
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F27	Problem Identified During Non-Clinical Procedure	A070504	Premature Discharge of Battery		
12434	2023	All other reportable incidents	47265	Dual-chamber implantable pacemaker, rate-responsive	CT141	Electrodes/Leads and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT145	Cardiovascular devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT233	Surgical	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT2406	Implantable	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT2407	Active implantable	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT2727	Active implant devices/ancillaries	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT2750	Implantable cardiac electrical stimulators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		

							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT2985	Active	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT3001	Implantable metal	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT3002	Implantable synthetic polymer	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT3006	Implantable inorganic material	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT314	Electrically powered	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT321	Long-term surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT449	Generators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT981	Single use	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT983	Surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
12435	2023		47265	Dual-chamber implantable pacemaker, rate-responsive	CT141	Electrodes/Leads and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	25/02/2023	25/02/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	25/02/2023	25/02/2023
					CT145	Cardiovascular devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	25/02/2023	25/02/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	25/02/2023	25/02/2023
					CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	25/02/2023	25/02/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	25/02/2023	25/02/2023
					CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	25/02/2023	25/02/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	25/02/2023	25/02/2023
					CT233	Surgical	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	25/02/2023	25/02/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	25/02/2023	25/02/2023
					CT2406	Implantable	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	25/02/2023	25/02/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	25/02/2023	25/02/2023
					CT2407	Active implantable	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	25/02/2023	25/02/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	25/02/2023	25/02/2023
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	25/02/2023	25/02/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	25/02/2023	25/02/2023
					CT2727	Active implant devices/ancillaries	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	25/02/2023	25/02/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	25/02/2023	25/02/2023
					CT2750	Implantable cardiac electrical stimulators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	25/02/2023	25/02/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	25/02/2023	25/02/2023
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	25/02/2023	25/02/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	25/02/2023	25/02/2023
					CT2985	Active	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	25/02/2023	25/02/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	25/02/2023	25/02/2023
					CT3001	Implantable metal	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	25/02/2023	25/02/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	25/02/2023	25/02/2023
					CT3002	Implantable synthetic polymer	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	25/02/2023	25/02/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	25/02/2023	25/02/2023
					CT3006	Implantable inorganic material	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	25/02/2023	25/02/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	25/02/2023	25/02/2023
					CT314	Electrically powered	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	25/02/2023	25/02/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	25/02/2023	25/02/2023
					CT321	Long-term surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	25/02/2023	25/02/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	25/02/2023	25/02/2023
					CT449	Generators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	25/02/2023	25/02/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	25/02/2023	25/02/2023
					CT981	Single use	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	25/02/2023	25/02/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	25/02/2023	25/02/2023
					CT983	Surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	25/02/2023	25/02/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	25/02/2023	25/02/2023
12436	2023		47265	Dual-chamber implantable pacemaker, rate-responsive	CT141	Electrodes/Leads and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	10/02/2023	10/02/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	10/02/2023	10/02/2023
					CT145	Cardiovascular devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	10/02/2023	10/02/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	10/02/2023	10/02/2023
					CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	10/02/2023	10/02/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	10/02/2023	10/02/2023
					CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	10/02/2023	10/02/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	10/02/2023	10/02/2023
					CT233	Surgical	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	10/02/2023	10/02/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	10/02/2023	10/02/2023
					CT2406	Implantable	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	10/02/2023	10/02/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	10/02/2023	10/02/2023
					CT2407	Active implantable	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	10/02/2023	10/02/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	10/02/2023	10/02/2023
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	10/02/2023	10/02/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	10/02/2023	10/02/2023
					CT2727	Active implant devices/ancillaries	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	10/02/2023	10/02/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	10/02/2023	10/02/2023
					CT2750	Implantable cardiac electrical stimulators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	10/02/2023	10/02/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	10/02/2023	10/02/2023
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	10/02/2023	10/02/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	10/02/2023	10/02/2023
					CT2985	Active	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	10/02/2023	10/02/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	10/02/2023	10/02/2023
					CT3001	Implantable metal	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	10/02/2023	10/02/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	10/02/2023	10/02/2023
					CT3002	Implantable synthetic polymer	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	10/02/2023	10/02/2023

							F1901	Additional Surgery	A24	No Apparent Device Problem	10/02/2023	10/02/2023
					CT3006	Implantable inorganic material	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	10/02/2023	10/02/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	10/02/2023	10/02/2023
					CT314	Electrically powered	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	10/02/2023	10/02/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	10/02/2023	10/02/2023
					CT321	Long-term surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	10/02/2023	10/02/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	10/02/2023	10/02/2023
					CT449	Generators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	10/02/2023	10/02/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	10/02/2023	10/02/2023
					CT981	Single use	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	10/02/2023	10/02/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	10/02/2023	10/02/2023
					CT983	Surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	10/02/2023	10/02/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	10/02/2023	10/02/2023
12437	2023		35224	External pacemaker, epicardial pacing	CT1684	Cardiac pacemakers and associated devices			A26	Insufficient Device Problem Information		
					CT1685	Cardiac pacemakers			A26	Insufficient Device Problem Information		
12438	2023	All other reportable incidents	35852	Single-chamber implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	16/03/2023	16/03/2023
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	16/03/2023	16/03/2023
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	16/03/2023	16/03/2023
12439	2023		47265	Dual-chamber implantable pacemaker, rate-responsive	CT141	Electrodes/Leads and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	26/01/2023	26/01/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	26/01/2023	26/01/2023
					CT145	Cardiovascular devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	26/01/2023	26/01/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	26/01/2023	26/01/2023
					CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	26/01/2023	26/01/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	26/01/2023	26/01/2023
					CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	26/01/2023	26/01/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	26/01/2023	26/01/2023
					CT233	Surgical	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	26/01/2023	26/01/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	26/01/2023	26/01/2023
					CT2406	Implantable	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	26/01/2023	26/01/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	26/01/2023	26/01/2023
					CT2407	Active implantable	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	26/01/2023	26/01/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	26/01/2023	26/01/2023
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	26/01/2023	26/01/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	26/01/2023	26/01/2023
					CT2727	Active implant devices/ancillaries	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	26/01/2023	26/01/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	26/01/2023	26/01/2023
					CT2750	Implantable cardiac electrical stimulators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	26/01/2023	26/01/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	26/01/2023	26/01/2023
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	26/01/2023	26/01/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	26/01/2023	26/01/2023
					CT2985	Active	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	26/01/2023	26/01/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	26/01/2023	26/01/2023
					CT3001	Implantable metal	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	26/01/2023	26/01/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	26/01/2023	26/01/2023
					CT3002	Implantable synthetic polymer	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	26/01/2023	26/01/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	26/01/2023	26/01/2023
					CT3006	Implantable inorganic material	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	26/01/2023	26/01/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	26/01/2023	26/01/2023
					CT314	Electrically powered	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	26/01/2023	26/01/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	26/01/2023	26/01/2023
					CT321	Long-term surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	26/01/2023	26/01/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	26/01/2023	26/01/2023
					CT449	Generators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	26/01/2023	26/01/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	26/01/2023	26/01/2023
					CT981	Single use	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	26/01/2023	26/01/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	26/01/2023	26/01/2023
					CT983	Surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	26/01/2023	26/01/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	26/01/2023	26/01/2023
12440	2023		47265	Dual-chamber implantable pacemaker, rate-responsive	CT141	Electrodes/Leads and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	17/02/2023	17/02/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	17/02/2023	17/02/2023
					CT145	Cardiovascular devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	17/02/2023	17/02/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	17/02/2023	17/02/2023
					CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	17/02/2023	17/02/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	17/02/2023	17/02/2023
					CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	17/02/2023	17/02/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	17/02/2023	17/02/2023
					CT233	Surgical	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	17/02/2023	17/02/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	17/02/2023	17/02/2023
					CT2406	Implantable	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	17/02/2023	17/02/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	17/02/2023	17/02/2023
					CT2407	Active implantable	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	17/02/2023	17/02/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	17/02/2023	17/02/2023
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	17/02/2023	17/02/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	17/02/2023	17/02/2023
					CT2727	Active implant devices/ancillaries	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	17/02/2023	17/02/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	17/02/2023	17/02/2023
					CT2750	Implantable cardiac electrical stimulators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	17/02/2023	17/02/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	17/02/2023	17/02/2023
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	17/02/2023	17/02/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	17/02/2023	17/02/2023
					CT2985	Active	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	17/02/2023	17/02/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	17/02/2023	17/02/2023
					CT3001	Implantable metal	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	17/02/2023	17/02/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	17/02/2023	17/02/2023
					CT3002	Implantable synthetic polymer	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	17/02/2023	17/02/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	17/02/2023	17/02/2023
					CT3006	Implantable inorganic material	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	17/02/2023	17/02/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	17/02/2023	17/02/2023

					CT314	Electrically powered	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	17/02/2023	17/02/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	17/02/2023	17/02/2023
					CT321	Long-term surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	17/02/2023	17/02/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	17/02/2023	17/02/2023
					CT449	Generators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	17/02/2023	17/02/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	17/02/2023	17/02/2023
					CT981	Single use	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	17/02/2023	17/02/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	17/02/2023	17/02/2023
					CT983	Surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	17/02/2023	17/02/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	17/02/2023	17/02/2023
12441	2023		47265	Dual-chamber implantable pacemaker, rate-responsive	CT141	Electrodes/Leads and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	22/02/2023	22/02/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	22/02/2023	22/02/2023
					CT145	Cardiovascular devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	22/02/2023	22/02/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	22/02/2023	22/02/2023
					CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	22/02/2023	22/02/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	22/02/2023	22/02/2023
					CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	22/02/2023	22/02/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	22/02/2023	22/02/2023
					CT233	Surgical	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	22/02/2023	22/02/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	22/02/2023	22/02/2023
					CT2406	Implantable	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	22/02/2023	22/02/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	22/02/2023	22/02/2023
					CT2407	Active implantable	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	22/02/2023	22/02/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	22/02/2023	22/02/2023
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	22/02/2023	22/02/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	22/02/2023	22/02/2023
					CT2727	Active implant devices/ancillaries	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	22/02/2023	22/02/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	22/02/2023	22/02/2023
					CT2750	Implantable cardiac electrical stimulators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	22/02/2023	22/02/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	22/02/2023	22/02/2023
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	22/02/2023	22/02/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	22/02/2023	22/02/2023
					CT2985	Active	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	22/02/2023	22/02/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	22/02/2023	22/02/2023
					CT3001	Implantable metal	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	22/02/2023	22/02/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	22/02/2023	22/02/2023
					CT3002	Implantable synthetic polymer	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	22/02/2023	22/02/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	22/02/2023	22/02/2023
					CT3006	Implantable inorganic material	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	22/02/2023	22/02/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	22/02/2023	22/02/2023
					CT314	Electrically powered	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	22/02/2023	22/02/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	22/02/2023	22/02/2023
					CT321	Long-term surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	22/02/2023	22/02/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	22/02/2023	22/02/2023
					CT449	Generators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	22/02/2023	22/02/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	22/02/2023	22/02/2023
					CT981	Single use	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	22/02/2023	22/02/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	22/02/2023	22/02/2023
					CT983	Surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	22/02/2023	22/02/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	22/02/2023	22/02/2023
12442	2023		47265	Dual-chamber implantable pacemaker, rate-responsive	CT141	Electrodes/Leads and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	15/02/2023	15/02/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	15/02/2023	15/02/2023
					CT145	Cardiovascular devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	15/02/2023	15/02/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	15/02/2023	15/02/2023
					CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	15/02/2023	15/02/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	15/02/2023	15/02/2023
					CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	15/02/2023	15/02/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	15/02/2023	15/02/2023
					CT233	Surgical	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	15/02/2023	15/02/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	15/02/2023	15/02/2023
					CT2406	Implantable	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	15/02/2023	15/02/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	15/02/2023	15/02/2023
					CT2407	Active implantable	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	15/02/2023	15/02/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	15/02/2023	15/02/2023
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	15/02/2023	15/02/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	15/02/2023	15/02/2023
					CT2727	Active implant devices/ancillaries	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	15/02/2023	15/02/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	15/02/2023	15/02/2023
					CT2750	Implantable cardiac electrical stimulators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	15/02/2023	15/02/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	15/02/2023	15/02/2023
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	15/02/2023	15/02/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	15/02/2023	15/02/2023
					CT2985	Active	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	15/02/2023	15/02/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	15/02/2023	15/02/2023
					CT3001	Implantable metal	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	15/02/2023	15/02/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	15/02/2023	15/02/2023
					CT3002	Implantable synthetic polymer	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	15/02/2023	15/02/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	15/02/2023	15/02/2023
					CT3006	Implantable inorganic material	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	15/02/2023	15/02/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	15/02/2023	15/02/2023
					CT314	Electrically powered	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	15/02/2023	15/02/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	15/02/2023	15/02/2023
					CT321	Long-term surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	15/02/2023	15/02/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	15/02/2023	15/02/2023
					CT449	Generators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	15/02/2023	15/02/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	15/02/2023	15/02/2023
					CT981	Single use	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	15/02/2023	15/02/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	15/02/2023	15/02/2023

					CT983	Surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	15/02/2023	15/02/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	15/02/2023	15/02/2023
12443	2023	All other reportable incidents	37265	Dual-chamber implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
12444	2023		47265	Dual-chamber implantable pacemaker, rate-responsive	CT141	Electrodes/Leads and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	13/02/2023	13/02/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	13/02/2023	13/02/2023
					CT145	Cardiovascular devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	13/02/2023	13/02/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	13/02/2023	13/02/2023
					CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	13/02/2023	13/02/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	13/02/2023	13/02/2023
					CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	13/02/2023	13/02/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	13/02/2023	13/02/2023
					CT233	Surgical	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	13/02/2023	13/02/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	13/02/2023	13/02/2023
					CT2406	Implantable	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	13/02/2023	13/02/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	13/02/2023	13/02/2023
					CT2407	Active implantable	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	13/02/2023	13/02/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	13/02/2023	13/02/2023
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	13/02/2023	13/02/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	13/02/2023	13/02/2023
					CT2727	Active implant devices/ancillaries	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	13/02/2023	13/02/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	13/02/2023	13/02/2023
					CT2750	Implantable cardiac electrical stimulators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	13/02/2023	13/02/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	13/02/2023	13/02/2023
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	13/02/2023	13/02/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	13/02/2023	13/02/2023
					CT2985	Active	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	13/02/2023	13/02/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	13/02/2023	13/02/2023
					CT3001	Implantable metal	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	13/02/2023	13/02/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	13/02/2023	13/02/2023
					CT3002	Implantable synthetic polymer	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	13/02/2023	13/02/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	13/02/2023	13/02/2023
					CT3006	Implantable inorganic material	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	13/02/2023	13/02/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	13/02/2023	13/02/2023
					CT314	Electrically powered	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	13/02/2023	13/02/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	13/02/2023	13/02/2023
					CT321	Long-term surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	13/02/2023	13/02/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	13/02/2023	13/02/2023
					CT449	Generators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	13/02/2023	13/02/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	13/02/2023	13/02/2023
					CT981	Single use	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	13/02/2023	13/02/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	13/02/2023	13/02/2023
					CT983	Surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	13/02/2023	13/02/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	13/02/2023	13/02/2023
12445	2023	All other reportable incidents	36241	Pacemaker/defibrillator lead	CT1684	Cardiac pacemakers and associated devices	E2104	Electric Shock	A070909	Oversensing		
									A071302	Inappropriate/Inadequate Shock/Stimulation		
									A072201	High Impedance		
									A090801	Signal Artifact/Noise		
							F08	Hospitalization or Prolonged Hospitalization	A070909	Oversensing		
									A071302	Inappropriate/Inadequate Shock/Stimulation		
									A072201	High Impedance		
									A090801	Signal Artifact/Noise		
							F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A070909	Oversensing		
									A071302	Inappropriate/Inadequate Shock/Stimulation		
									A072201	High Impedance		
									A090801	Signal Artifact/Noise		
							F1904	Device Repositioning	A070909	Oversensing		
									A071302	Inappropriate/Inadequate Shock/Stimulation		
									A072201	High Impedance		
									A090801	Signal Artifact/Noise		
12446	2023	All other reportable incidents	47265	Dual-chamber implantable pacemaker, rate-responsive	CT141	Electrodes/Leads and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070910	Undersensing		
							F15	Recognized Device or Procedural Complication	A070910	Undersensing		
					CT145	Cardiovascular devices	E2403	No Clinical Signs, Symptoms or Conditions	A070910	Undersensing		
							F15	Recognized Device or Procedural Complication	A070910	Undersensing		
					CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070910	Undersensing		
							F15	Recognized Device or Procedural Complication	A070910	Undersensing		
					CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A070910	Undersensing		
							F15	Recognized Device or Procedural Complication	A070910	Undersensing		
					CT233	Surgical	E2403	No Clinical Signs, Symptoms or Conditions	A070910	Undersensing		
							F15	Recognized Device or Procedural Complication	A070910	Undersensing		
					CT2406	Implantable	E2403	No Clinical Signs, Symptoms or Conditions	A070910	Undersensing		
							F15	Recognized Device or Procedural Complication	A070910	Undersensing		
					CT2407	Active implantable	E2403	No Clinical Signs, Symptoms or Conditions	A070910	Undersensing		
							F15	Recognized Device or Procedural Complication	A070910	Undersensing		
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A070910	Undersensing		
							F15	Recognized Device or Procedural Complication	A070910	Undersensing		
					CT2727	Active implant devices/ancillaries	E2403	No Clinical Signs, Symptoms or Conditions	A070910	Undersensing		
							F15	Recognized Device or Procedural Complication	A070910	Undersensing		
					CT2750	Implantable cardiac electrical stimulators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070910	Undersensing		
							F15	Recognized Device or Procedural Complication	A070910	Undersensing		
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A070910	Undersensing		
							F15	Recognized Device or Procedural Complication	A070910	Undersensing		
					CT2985	Active	E2403	No Clinical Signs, Symptoms or Conditions	A070910	Undersensing		
							F15	Recognized Device or Procedural Complication	A070910	Undersensing		
					CT3001	Implantable metal	E2403	No Clinical Signs, Symptoms or Conditions	A070910	Undersensing		
							F15	Recognized Device or Procedural Complication	A070910	Undersensing		

					CT3002	Implantable synthetic polymer	E2403	No Clinical Signs, Symptoms or Conditions	A070910	Undersensing		
							F15	Recognized Device or Procedural Complication	A070910	Undersensing		
					CT3006	Implantable inorganic material	E2403	No Clinical Signs, Symptoms or Conditions	A070910	Undersensing		
							F15	Recognized Device or Procedural Complication	A070910	Undersensing		
					CT314	Electrically powered	E2403	No Clinical Signs, Symptoms or Conditions	A070910	Undersensing		
							F15	Recognized Device or Procedural Complication	A070910	Undersensing		
					CT321	Long-term surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A070910	Undersensing		
							F15	Recognized Device or Procedural Complication	A070910	Undersensing		
					CT449	Generators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070910	Undersensing		
							F15	Recognized Device or Procedural Complication	A070910	Undersensing		
					CT981	Single use	E2403	No Clinical Signs, Symptoms or Conditions	A070910	Undersensing		
							F15	Recognized Device or Procedural Complication	A070910	Undersensing		
					CT983	Surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A070910	Undersensing		
							F15	Recognized Device or Procedural Complication	A070910	Undersensing		
12447	2023	All other reportable incidents	47265	Dual-chamber implantable pacemaker, rate-responsive	CT141	Electrodes/Leads and associated devices	E0112	Dizziness	A090201	Device Displays Incorrect Message		
									A090801	Signal Artifact/Noise		
							E011903	Syncope/Fainting	A090201	Device Displays Incorrect Message		
									A090801	Signal Artifact/Noise		
							F08	Hospitalization or Prolonged Hospitalization	A090201	Device Displays Incorrect Message		
									A090801	Signal Artifact/Noise		
					CT145	Cardiovascular devices	E0112	Dizziness	A090201	Device Displays Incorrect Message		
									A090801	Signal Artifact/Noise		
							E011903	Syncope/Fainting	A090201	Device Displays Incorrect Message		
									A090801	Signal Artifact/Noise		
							F08	Hospitalization or Prolonged Hospitalization	A090201	Device Displays Incorrect Message		
									A090801	Signal Artifact/Noise		
					CT1684	Cardiac pacemakers and associated devices	E0112	Dizziness	A090201	Device Displays Incorrect Message		
									A090801	Signal Artifact/Noise		
							E011903	Syncope/Fainting	A090201	Device Displays Incorrect Message		
									A090801	Signal Artifact/Noise		
							F08	Hospitalization or Prolonged Hospitalization	A090201	Device Displays Incorrect Message		
									A090801	Signal Artifact/Noise		
					CT1685	Cardiac pacemakers	E0112	Dizziness	A090201	Device Displays Incorrect Message		
									A090801	Signal Artifact/Noise		
							E011903	Syncope/Fainting	A090201	Device Displays Incorrect Message		
									A090801	Signal Artifact/Noise		
							F08	Hospitalization or Prolonged Hospitalization	A090201	Device Displays Incorrect Message		
									A090801	Signal Artifact/Noise		
					CT233	Surgical	E0112	Dizziness	A090201	Device Displays Incorrect Message		
									A090801	Signal Artifact/Noise		
							E011903	Syncope/Fainting	A090201	Device Displays Incorrect Message		
									A090801	Signal Artifact/Noise		
							F08	Hospitalization or Prolonged Hospitalization	A090201	Device Displays Incorrect Message		
									A090801	Signal Artifact/Noise		
					CT2406	Implantable	E0112	Dizziness	A090201	Device Displays Incorrect Message		
									A090801	Signal Artifact/Noise		
							E011903	Syncope/Fainting	A090201	Device Displays Incorrect Message		
									A090801	Signal Artifact/Noise		
							F08	Hospitalization or Prolonged Hospitalization	A090201	Device Displays Incorrect Message		
									A090801	Signal Artifact/Noise		
					CT2407	Active implantable	E0112	Dizziness	A090201	Device Displays Incorrect Message		
									A090801	Signal Artifact/Noise		
							E011903	Syncope/Fainting	A090201	Device Displays Incorrect Message		
									A090801	Signal Artifact/Noise		
							F08	Hospitalization or Prolonged Hospitalization	A090201	Device Displays Incorrect Message		
									A090801	Signal Artifact/Noise		
					CT257	Implantable cardiac pacemakers	E0112	Dizziness	A090201	Device Displays Incorrect Message		
									A090801	Signal Artifact/Noise		
							E011903	Syncope/Fainting	A090201	Device Displays Incorrect Message		
									A090801	Signal Artifact/Noise		
							F08	Hospitalization or Prolonged Hospitalization	A090201	Device Displays Incorrect Message		
									A090801	Signal Artifact/Noise		
					CT2727	Active implant devices/ancillaries	E0112	Dizziness	A090201	Device Displays Incorrect Message		
									A090801	Signal Artifact/Noise		
							E011903	Syncope/Fainting	A090201	Device Displays Incorrect Message		
									A090801	Signal Artifact/Noise		
							F08	Hospitalization or Prolonged Hospitalization	A090201	Device Displays Incorrect Message		
									A090801	Signal Artifact/Noise		
					CT2750	Implantable cardiac electrical stimulators and associated devices	E0112	Dizziness	A090201	Device Displays Incorrect Message		
									A090801	Signal Artifact/Noise		
							E011903	Syncope/Fainting	A090201	Device Displays Incorrect Message		
									A090801	Signal Artifact/Noise		
							F08	Hospitalization or Prolonged Hospitalization	A090201	Device Displays Incorrect Message		
									A090801	Signal Artifact/Noise		
					CT2866	Implantable cardiac electrical stimulators	E0112	Dizziness	A090201	Device Displays Incorrect Message		
									A090801	Signal Artifact/Noise		
							E011903	Syncope/Fainting	A090201	Device Displays Incorrect Message		
									A090801	Signal Artifact/Noise		
							F08	Hospitalization or Prolonged Hospitalization	A090201	Device Displays Incorrect Message		
									A090801	Signal Artifact/Noise		
					CT2985	Active	E0112	Dizziness	A090201	Device Displays Incorrect Message		
									A090801	Signal Artifact/Noise		
							E011903	Syncope/Fainting	A090201	Device Displays Incorrect Message		
									A090801	Signal Artifact/Noise		
							F08	Hospitalization or Prolonged Hospitalization	A090201	Device Displays Incorrect Message		
									A090801	Signal Artifact/Noise		
					CT3001	Implantable metal	E0112	Dizziness	A090201	Device Displays Incorrect Message		
									A090801	Signal Artifact/Noise		

							E011903	Syncope/Fainting	A090201	Device Displays Incorrect Message		
									A090801	Signal Artifact/Noise		
							F08	Hospitalization or Prolonged Hospitalization	A090201	Device Displays Incorrect Message		
									A090801	Signal Artifact/Noise		
					CT3002	Implantable synthetic polymer	E0112	Dizziness	A090201	Device Displays Incorrect Message		
									A090801	Signal Artifact/Noise		
							E011903	Syncope/Fainting	A090201	Device Displays Incorrect Message		
									A090801	Signal Artifact/Noise		
							F08	Hospitalization or Prolonged Hospitalization	A090201	Device Displays Incorrect Message		
									A090801	Signal Artifact/Noise		
					CT3006	Implantable inorganic material	E0112	Dizziness	A090201	Device Displays Incorrect Message		
									A090801	Signal Artifact/Noise		
							E011903	Syncope/Fainting	A090201	Device Displays Incorrect Message		
									A090801	Signal Artifact/Noise		
							F08	Hospitalization or Prolonged Hospitalization	A090201	Device Displays Incorrect Message		
									A090801	Signal Artifact/Noise		
					CT314	Electrically powered	E0112	Dizziness	A090201	Device Displays Incorrect Message		
									A090801	Signal Artifact/Noise		
							E011903	Syncope/Fainting	A090201	Device Displays Incorrect Message		
									A090801	Signal Artifact/Noise		
							F08	Hospitalization or Prolonged Hospitalization	A090201	Device Displays Incorrect Message		
									A090801	Signal Artifact/Noise		
					CT321	Long-term surgical invasive	E0112	Dizziness	A090201	Device Displays Incorrect Message		
									A090801	Signal Artifact/Noise		
							E011903	Syncope/Fainting	A090201	Device Displays Incorrect Message		
									A090801	Signal Artifact/Noise		
							F08	Hospitalization or Prolonged Hospitalization	A090201	Device Displays Incorrect Message		
									A090801	Signal Artifact/Noise		
					CT449	Generators and associated devices	E0112	Dizziness	A090201	Device Displays Incorrect Message		
									A090801	Signal Artifact/Noise		
							E011903	Syncope/Fainting	A090201	Device Displays Incorrect Message		
									A090801	Signal Artifact/Noise		
							F08	Hospitalization or Prolonged Hospitalization	A090201	Device Displays Incorrect Message		
									A090801	Signal Artifact/Noise		
					CT981	Single use	E0112	Dizziness	A090201	Device Displays Incorrect Message		
									A090801	Signal Artifact/Noise		
							E011903	Syncope/Fainting	A090201	Device Displays Incorrect Message		
									A090801	Signal Artifact/Noise		
							F08	Hospitalization or Prolonged Hospitalization	A090201	Device Displays Incorrect Message		
									A090801	Signal Artifact/Noise		
					CT983	Surgical invasive	E0112	Dizziness	A090201	Device Displays Incorrect Message		
									A090801	Signal Artifact/Noise		
							E011903	Syncope/Fainting	A090201	Device Displays Incorrect Message		
									A090801	Signal Artifact/Noise		
							F08	Hospitalization or Prolonged Hospitalization	A090201	Device Displays Incorrect Message		
									A090801	Signal Artifact/Noise		
12448	2023		47265	Dual-chamber implantable pacemaker, rate-responsive	CT141	Electrodes/Leads and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	23/02/2023	23/02/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	23/02/2023	23/02/2023
					CT145	Cardiovascular devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	23/02/2023	23/02/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	23/02/2023	23/02/2023
					CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	23/02/2023	23/02/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	23/02/2023	23/02/2023
					CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	23/02/2023	23/02/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	23/02/2023	23/02/2023
					CT233	Surgical	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	23/02/2023	23/02/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	23/02/2023	23/02/2023
					CT2406	Implantable	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	23/02/2023	23/02/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	23/02/2023	23/02/2023
					CT2407	Active implantable	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	23/02/2023	23/02/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	23/02/2023	23/02/2023
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	23/02/2023	23/02/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	23/02/2023	23/02/2023
					CT2727	Active implant devices/ancillaries	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	23/02/2023	23/02/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	23/02/2023	23/02/2023
					CT2750	Implantable cardiac electrical stimulators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	23/02/2023	23/02/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	23/02/2023	23/02/2023
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	23/02/2023	23/02/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	23/02/2023	23/02/2023
					CT2985	Active	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	23/02/2023	23/02/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	23/02/2023	23/02/2023
					CT3001	Implantable metal	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	23/02/2023	23/02/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	23/02/2023	23/02/2023
					CT3002	Implantable synthetic polymer	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	23/02/2023	23/02/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	23/02/2023	23/02/2023
					CT3006	Implantable inorganic material	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	23/02/2023	23/02/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	23/02/2023	23/02/2023
					CT314	Electrically powered	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	23/02/2023	23/02/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	23/02/2023	23/02/2023
					CT321	Long-term surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	23/02/2023	23/02/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	23/02/2023	23/02/2023
					CT449	Generators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	23/02/2023	23/02/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	23/02/2023	23/02/2023
					CT981	Single use	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	23/02/2023	23/02/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	23/02/2023	23/02/2023
					CT983	Surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	23/02/2023	23/02/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	23/02/2023	23/02/2023
12449	2023	All other reportable incidents	35852	Single-chamber implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		

							F26	No Health Consequences or Impact	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
12450	2023	All other reportable incidents	17882	Physiologic-monitoring defibrillation system	CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A0719	Unexpected Shutdown		
							F27	Problem Identified During Non-Clinical Procedure	A0719	Unexpected Shutdown		
12451	2023		47265	Dual-chamber implantable pacemaker, rate-responsive	CT141	Electrodes/Leads and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	20/03/2023	20/03/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	20/03/2023	20/03/2023
					CT145	Cardiovascular devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	20/03/2023	20/03/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	20/03/2023	20/03/2023
					CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	20/03/2023	20/03/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	20/03/2023	20/03/2023
					CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	20/03/2023	20/03/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	20/03/2023	20/03/2023
					CT233	Surgical	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	20/03/2023	20/03/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	20/03/2023	20/03/2023
					CT2406	Implantable	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	20/03/2023	20/03/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	20/03/2023	20/03/2023
					CT2407	Active implantable	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	20/03/2023	20/03/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	20/03/2023	20/03/2023
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	20/03/2023	20/03/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	20/03/2023	20/03/2023
					CT2727	Active implant devices/ancillaries	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	20/03/2023	20/03/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	20/03/2023	20/03/2023
					CT2750	Implantable cardiac electrical stimulators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	20/03/2023	20/03/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	20/03/2023	20/03/2023
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	20/03/2023	20/03/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	20/03/2023	20/03/2023
					CT2985	Active	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	20/03/2023	20/03/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	20/03/2023	20/03/2023
					CT3001	Implantable metal	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	20/03/2023	20/03/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	20/03/2023	20/03/2023
					CT3002	Implantable synthetic polymer	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	20/03/2023	20/03/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	20/03/2023	20/03/2023
					CT3006	Implantable inorganic material	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	20/03/2023	20/03/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	20/03/2023	20/03/2023
					CT314	Electrically powered	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	20/03/2023	20/03/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	20/03/2023	20/03/2023
					CT321	Long-term surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	20/03/2023	20/03/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	20/03/2023	20/03/2023
					CT449	Generators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	20/03/2023	20/03/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	20/03/2023	20/03/2023
					CT981	Single use	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	20/03/2023	20/03/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	20/03/2023	20/03/2023
					CT983	Surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	20/03/2023	20/03/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	20/03/2023	20/03/2023
12452	2023	All other reportable incidents	17882	Physiologic-monitoring defibrillation system	CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A0719	Unexpected Shutdown		
							F27	Problem Identified During Non-Clinical Procedure	A0719	Unexpected Shutdown		
12453	2023	All other reportable incidents	47265	Dual-chamber implantable pacemaker, rate-responsive	CT141	Electrodes/Leads and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	01/02/2023	01/02/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	01/02/2023	01/02/2023
					CT145	Cardiovascular devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	01/02/2023	01/02/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	01/02/2023	01/02/2023
					CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	01/02/2023	01/02/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	01/02/2023	01/02/2023
					CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	01/02/2023	01/02/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	01/02/2023	01/02/2023
					CT233	Surgical	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	01/02/2023	01/02/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	01/02/2023	01/02/2023
					CT2406	Implantable	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	01/02/2023	01/02/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	01/02/2023	01/02/2023
					CT2407	Active implantable	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	01/02/2023	01/02/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	01/02/2023	01/02/2023
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	01/02/2023	01/02/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	01/02/2023	01/02/2023
					CT2727	Active implant devices/ancillaries	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	01/02/2023	01/02/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	01/02/2023	01/02/2023
					CT2750	Implantable cardiac electrical stimulators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	01/02/2023	01/02/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	01/02/2023	01/02/2023
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	01/02/2023	01/02/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	01/02/2023	01/02/2023
					CT2985	Active	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	01/02/2023	01/02/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	01/02/2023	01/02/2023
					CT3001	Implantable metal	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	01/02/2023	01/02/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	01/02/2023	01/02/2023
					CT3002	Implantable synthetic polymer	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	01/02/2023	01/02/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	01/02/2023	01/02/2023
					CT3006	Implantable inorganic material	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	01/02/2023	01/02/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	01/02/2023	01/02/2023
					CT314	Electrically powered	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	01/02/2023	01/02/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	01/02/2023	01/02/2023
					CT321	Long-term surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	01/02/2023	01/02/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	01/02/2023	01/02/2023
					CT449	Generators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	01/02/2023	01/02/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	01/02/2023	01/02/2023
					CT981	Single use	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	01/02/2023	01/02/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	01/02/2023	01/02/2023
					CT983	Surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	01/02/2023	01/02/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	01/02/2023	01/02/2023
12454	2023	All other reportable incidents	35852	Single-chamber implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	04/07/2023	04/07/2023

									A0908	Incorrect, Inadequate or Imprecise Result or Readings	04/07/2023	04/07/2023
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	04/07/2023	04/07/2023
									A0908	Incorrect, Inadequate or Imprecise Result or Readings	04/07/2023	04/07/2023
							F1903	Device Explantation	A070504	Premature Discharge of Battery	04/07/2023	04/07/2023
									A0908	Incorrect, Inadequate or Imprecise Result or Readings	04/07/2023	04/07/2023
12455	2023	All other reportable incidents	36241	Pacemaker/defibrillator lead	CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070101	Failure to Capture		
							F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A070101	Failure to Capture		
12456	2023	All other reportable incidents	36241	Pacemaker/defibrillator lead	CT1684	Cardiac pacemakers and associated devices	E0515	Vascular Dissection	A150206	Difficult to Insert	14/02/2023	14/02/2023
							E0619	Pericardial Effusion	A150206	Difficult to Insert	14/02/2023	14/02/2023
							F1908	Prolonged Surgery	A150206	Difficult to Insert	14/02/2023	14/02/2023
							F2203	Imaging Required	A150206	Difficult to Insert	14/02/2023	14/02/2023
12457	2023	All other reportable incidents	35852	Single-chamber implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E2104	Electric Shock	A070909	Oversensing		
									A071302	Inappropriate/Inadequate Shock/Stimulation		
									A090801	Signal Artifact/Noise		
									A1301	Failure to Read Input Signal		
							F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A070909	Oversensing		
									A071302	Inappropriate/Inadequate Shock/Stimulation		
									A090801	Signal Artifact/Noise		
									A1301	Failure to Read Input Signal		
							F2203	Imaging Required	A070909	Oversensing		
									A071302	Inappropriate/Inadequate Shock/Stimulation		
									A090801	Signal Artifact/Noise		
									A1301	Failure to Read Input Signal		
12458	2023	Death	60789	Intracardiac pacemaker	CT130	Catheters and associated devices	E0109	Convulsion/Seizure	A24	No Apparent Device Problem		
							E0306	Sepsis	A24	No Apparent Device Problem		
							E0602	Cardiac Arrest	A24	No Apparent Device Problem		
							E0619	Pericardial Effusion	A24	No Apparent Device Problem		
							E0743	Respiratory Insufficiency	A24	No Apparent Device Problem		
							E233605	Septic Shock	A24	No Apparent Device Problem		
							F02	Death	A24	No Apparent Device Problem		
							F23	Unexpected Medical Intervention	A24	No Apparent Device Problem		
							F2303	Medication Required	A24	No Apparent Device Problem		
					CT145	Cardiovascular devices	E0109	Convulsion/Seizure	A24	No Apparent Device Problem		
							E0306	Sepsis	A24	No Apparent Device Problem		
							E0602	Cardiac Arrest	A24	No Apparent Device Problem		
							E0619	Pericardial Effusion	A24	No Apparent Device Problem		
							E0743	Respiratory Insufficiency	A24	No Apparent Device Problem		
							E233605	Septic Shock	A24	No Apparent Device Problem		
							F02	Death	A24	No Apparent Device Problem		
							F23	Unexpected Medical Intervention	A24	No Apparent Device Problem		
							F2303	Medication Required	A24	No Apparent Device Problem		
					CT1684	Cardiac pacemakers and associated devices	E0109	Convulsion/Seizure	A24	No Apparent Device Problem		
							E0306	Sepsis	A24	No Apparent Device Problem		
							E0602	Cardiac Arrest	A24	No Apparent Device Problem		
							E0619	Pericardial Effusion	A24	No Apparent Device Problem		
							E0743	Respiratory Insufficiency	A24	No Apparent Device Problem		
							E233605	Septic Shock	A24	No Apparent Device Problem		
							F02	Death	A24	No Apparent Device Problem		
							F23	Unexpected Medical Intervention	A24	No Apparent Device Problem		
							F2303	Medication Required	A24	No Apparent Device Problem		
					CT1685	Cardiac pacemakers	E0109	Convulsion/Seizure	A24	No Apparent Device Problem		
							E0306	Sepsis	A24	No Apparent Device Problem		
							E0602	Cardiac Arrest	A24	No Apparent Device Problem		
							E0619	Pericardial Effusion	A24	No Apparent Device Problem		
							E0743	Respiratory Insufficiency	A24	No Apparent Device Problem		
							E233605	Septic Shock	A24	No Apparent Device Problem		
							F02	Death	A24	No Apparent Device Problem		
							F23	Unexpected Medical Intervention	A24	No Apparent Device Problem		
							F2303	Medication Required	A24	No Apparent Device Problem		
					CT233	Surgical	E0109	Convulsion/Seizure	A24	No Apparent Device Problem		
							E0306	Sepsis	A24	No Apparent Device Problem		
							E0602	Cardiac Arrest	A24	No Apparent Device Problem		
							E0619	Pericardial Effusion	A24	No Apparent Device Problem		
							E0743	Respiratory Insufficiency	A24	No Apparent Device Problem		
							E233605	Septic Shock	A24	No Apparent Device Problem		
							F02	Death	A24	No Apparent Device Problem		
							F23	Unexpected Medical Intervention	A24	No Apparent Device Problem		
							F2303	Medication Required	A24	No Apparent Device Problem		
					CT2406	Implantable	E0109	Convulsion/Seizure	A24	No Apparent Device Problem		
							E0306	Sepsis	A24	No Apparent Device Problem		
							E0602	Cardiac Arrest	A24	No Apparent Device Problem		
							E0619	Pericardial Effusion	A24	No Apparent Device Problem		
							E0743	Respiratory Insufficiency	A24	No Apparent Device Problem		
							E233605	Septic Shock	A24	No Apparent Device Problem		
							F02	Death	A24	No Apparent Device Problem		
							F23	Unexpected Medical Intervention	A24	No Apparent Device Problem		
							F2303	Medication Required	A24	No Apparent Device Problem		
					CT2407	Active implantable	E0109	Convulsion/Seizure	A24	No Apparent Device Problem		
							E0306	Sepsis	A24	No Apparent Device Problem		
							E0602	Cardiac Arrest	A24	No Apparent Device Problem		
							E0619	Pericardial Effusion	A24	No Apparent Device Problem		
							E0743	Respiratory Insufficiency	A24	No Apparent Device Problem		
							E233605	Septic Shock	A24	No Apparent Device Problem		
							F02	Death	A24	No Apparent Device Problem		
							F23	Unexpected Medical Intervention	A24	No Apparent Device Problem		

							F2303	Medication Required	A24	No Apparent Device Problem		
					CT257	Implantable cardiac pacemakers	E0109	Convulsion/Seizure	A24	No Apparent Device Problem		
							E0306	Sepsis	A24	No Apparent Device Problem		
							E0602	Cardiac Arrest	A24	No Apparent Device Problem		
							E0619	Pericardial Effusion	A24	No Apparent Device Problem		
							E0743	Respiratory Insufficiency	A24	No Apparent Device Problem		
							E233605	Septic Shock	A24	No Apparent Device Problem		
							F02	Death	A24	No Apparent Device Problem		
							F23	Unexpected Medical Intervention	A24	No Apparent Device Problem		
							F2303	Medication Required	A24	No Apparent Device Problem		
					CT2727	Active implant devices/ancillaries	E0109	Convulsion/Seizure	A24	No Apparent Device Problem		
							E0306	Sepsis	A24	No Apparent Device Problem		
							E0602	Cardiac Arrest	A24	No Apparent Device Problem		
							E0619	Pericardial Effusion	A24	No Apparent Device Problem		
							E0743	Respiratory Insufficiency	A24	No Apparent Device Problem		
							E233605	Septic Shock	A24	No Apparent Device Problem		
							F02	Death	A24	No Apparent Device Problem		
							F23	Unexpected Medical Intervention	A24	No Apparent Device Problem		
							F2303	Medication Required	A24	No Apparent Device Problem		
					CT2750	Implantable cardiac electrical stimulators and associated devices	E0109	Convulsion/Seizure	A24	No Apparent Device Problem		
							E0306	Sepsis	A24	No Apparent Device Problem		
							E0602	Cardiac Arrest	A24	No Apparent Device Problem		
							E0619	Pericardial Effusion	A24	No Apparent Device Problem		
							E0743	Respiratory Insufficiency	A24	No Apparent Device Problem		
							E233605	Septic Shock	A24	No Apparent Device Problem		
							F02	Death	A24	No Apparent Device Problem		
							F23	Unexpected Medical Intervention	A24	No Apparent Device Problem		
							F2303	Medication Required	A24	No Apparent Device Problem		
					CT2818	Catheterization/Cannulation/Fluid path devices	E0109	Convulsion/Seizure	A24	No Apparent Device Problem		
							E0306	Sepsis	A24	No Apparent Device Problem		
							E0602	Cardiac Arrest	A24	No Apparent Device Problem		
							E0619	Pericardial Effusion	A24	No Apparent Device Problem		
							E0743	Respiratory Insufficiency	A24	No Apparent Device Problem		
							E233605	Septic Shock	A24	No Apparent Device Problem		
							F02	Death	A24	No Apparent Device Problem		
							F23	Unexpected Medical Intervention	A24	No Apparent Device Problem		
							F2303	Medication Required	A24	No Apparent Device Problem		
					CT2866	Implantable cardiac electrical stimulators	E0109	Convulsion/Seizure	A24	No Apparent Device Problem		
							E0306	Sepsis	A24	No Apparent Device Problem		
							E0602	Cardiac Arrest	A24	No Apparent Device Problem		
							E0619	Pericardial Effusion	A24	No Apparent Device Problem		
							E0743	Respiratory Insufficiency	A24	No Apparent Device Problem		
							E233605	Septic Shock	A24	No Apparent Device Problem		
							F02	Death	A24	No Apparent Device Problem		
							F23	Unexpected Medical Intervention	A24	No Apparent Device Problem		
							F2303	Medication Required	A24	No Apparent Device Problem		
					CT2985	Active	E0109	Convulsion/Seizure	A24	No Apparent Device Problem		
							E0306	Sepsis	A24	No Apparent Device Problem		
							E0602	Cardiac Arrest	A24	No Apparent Device Problem		
							E0619	Pericardial Effusion	A24	No Apparent Device Problem		
							E0743	Respiratory Insufficiency	A24	No Apparent Device Problem		
							E233605	Septic Shock	A24	No Apparent Device Problem		
							F02	Death	A24	No Apparent Device Problem		
							F23	Unexpected Medical Intervention	A24	No Apparent Device Problem		
							F2303	Medication Required	A24	No Apparent Device Problem		
					CT3001	Implantable metal	E0109	Convulsion/Seizure	A24	No Apparent Device Problem		
							E0306	Sepsis	A24	No Apparent Device Problem		
							E0602	Cardiac Arrest	A24	No Apparent Device Problem		
							E0619	Pericardial Effusion	A24	No Apparent Device Problem		
							E0743	Respiratory Insufficiency	A24	No Apparent Device Problem		
							E233605	Septic Shock	A24	No Apparent Device Problem		
							F02	Death	A24	No Apparent Device Problem		
							F23	Unexpected Medical Intervention	A24	No Apparent Device Problem		
							F2303	Medication Required	A24	No Apparent Device Problem		
					CT3002	Implantable synthetic polymer	E0109	Convulsion/Seizure	A24	No Apparent Device Problem		
							E0306	Sepsis	A24	No Apparent Device Problem		
							E0602	Cardiac Arrest	A24	No Apparent Device Problem		
							E0619	Pericardial Effusion	A24	No Apparent Device Problem		
							E0743	Respiratory Insufficiency	A24	No Apparent Device Problem		
							E233605	Septic Shock	A24	No Apparent Device Problem		
							F02	Death	A24	No Apparent Device Problem		
							F23	Unexpected Medical Intervention	A24	No Apparent Device Problem		
							F2303	Medication Required	A24	No Apparent Device Problem		
					CT3006	Implantable inorganic material	E0109	Convulsion/Seizure	A24	No Apparent Device Problem		
							E0306	Sepsis	A24	No Apparent Device Problem		
							E0602	Cardiac Arrest	A24	No Apparent Device Problem		
							E0619	Pericardial Effusion	A24	No Apparent Device Problem		
							E0743	Respiratory Insufficiency	A24	No Apparent Device Problem		
							E233605	Septic Shock	A24	No Apparent Device Problem		
							F02	Death	A24	No Apparent Device Problem		
							F23	Unexpected Medical Intervention	A24	No Apparent Device Problem		
							F2303	Medication Required	A24	No Apparent Device Problem		
					CT314	Electrically powered	E0109	Convulsion/Seizure	A24	No Apparent Device Problem		
							E0306	Sepsis	A24	No Apparent Device Problem		
							E0602	Cardiac Arrest	A24	No Apparent Device Problem		
							E0619	Pericardial Effusion	A24	No Apparent Device Problem		
							E0743	Respiratory Insufficiency	A24	No Apparent Device Problem		
							E233605	Septic Shock	A24	No Apparent Device Problem		

							F02	Death	A24	No Apparent Device Problem		
							F23	Unexpected Medical Intervention	A24	No Apparent Device Problem		
							F2303	Medication Required	A24	No Apparent Device Problem		
					CT321	Long-term surgical invasive	E0109	Convulsion/Seizure	A24	No Apparent Device Problem		
							E0306	Sepsis	A24	No Apparent Device Problem		
							E0602	Cardiac Arrest	A24	No Apparent Device Problem		
							E0619	Pericardial Effusion	A24	No Apparent Device Problem		
							E0743	Respiratory Insufficiency	A24	No Apparent Device Problem		
							E233605	Septic Shock	A24	No Apparent Device Problem		
							F02	Death	A24	No Apparent Device Problem		
							F23	Unexpected Medical Intervention	A24	No Apparent Device Problem		
							F2303	Medication Required	A24	No Apparent Device Problem		
					CT981	Single use	E0109	Convulsion/Seizure	A24	No Apparent Device Problem		
							E0306	Sepsis	A24	No Apparent Device Problem		
							E0602	Cardiac Arrest	A24	No Apparent Device Problem		
							E0619	Pericardial Effusion	A24	No Apparent Device Problem		
							E0743	Respiratory Insufficiency	A24	No Apparent Device Problem		
							E233605	Septic Shock	A24	No Apparent Device Problem		
							F02	Death	A24	No Apparent Device Problem		
							F23	Unexpected Medical Intervention	A24	No Apparent Device Problem		
							F2303	Medication Required	A24	No Apparent Device Problem		
					CT983	Surgical invasive	E0109	Convulsion/Seizure	A24	No Apparent Device Problem		
							E0306	Sepsis	A24	No Apparent Device Problem		
							E0602	Cardiac Arrest	A24	No Apparent Device Problem		
							E0619	Pericardial Effusion	A24	No Apparent Device Problem		
							E0743	Respiratory Insufficiency	A24	No Apparent Device Problem		
							E233605	Septic Shock	A24	No Apparent Device Problem		
							F02	Death	A24	No Apparent Device Problem		
							F23	Unexpected Medical Intervention	A24	No Apparent Device Problem		
							F2303	Medication Required	A24	No Apparent Device Problem		
12459	2023	All other reportable incidents	47265	Dual-chamber implantable pacemaker, rate-responsive	CT141	Electrodes/Leads and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	08/02/2023	08/02/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	08/02/2023	08/02/2023
					CT145	Cardiovascular devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	08/02/2023	08/02/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	08/02/2023	08/02/2023
					CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	08/02/2023	08/02/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	08/02/2023	08/02/2023
					CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	08/02/2023	08/02/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	08/02/2023	08/02/2023
					CT233	Surgical	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	08/02/2023	08/02/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	08/02/2023	08/02/2023
					CT2406	Implantable	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	08/02/2023	08/02/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	08/02/2023	08/02/2023
					CT2407	Active implantable	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	08/02/2023	08/02/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	08/02/2023	08/02/2023
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	08/02/2023	08/02/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	08/02/2023	08/02/2023
					CT2727	Active implant devices/ancillaries	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	08/02/2023	08/02/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	08/02/2023	08/02/2023
					CT2750	Implantable cardiac electrical stimulators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	08/02/2023	08/02/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	08/02/2023	08/02/2023
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	08/02/2023	08/02/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	08/02/2023	08/02/2023
					CT2985	Active	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	08/02/2023	08/02/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	08/02/2023	08/02/2023
					CT3001	Implantable metal	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	08/02/2023	08/02/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	08/02/2023	08/02/2023
					CT3002	Implantable synthetic polymer	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	08/02/2023	08/02/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	08/02/2023	08/02/2023
					CT3006	Implantable inorganic material	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	08/02/2023	08/02/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	08/02/2023	08/02/2023
					CT314	Electrically powered	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	08/02/2023	08/02/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	08/02/2023	08/02/2023
					CT321	Long-term surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	08/02/2023	08/02/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	08/02/2023	08/02/2023
					CT449	Generators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	08/02/2023	08/02/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	08/02/2023	08/02/2023
					CT981	Single use	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	08/02/2023	08/02/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	08/02/2023	08/02/2023
					CT983	Surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	08/02/2023	08/02/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	08/02/2023	08/02/2023
12460	2023	All other reportable incidents	17882	Physiologic-monitoring defibrillation system	CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070803	Failure to Power Up		
							F27	Problem Identified During Non-Clinical Procedure	A070803	Failure to Power Up		
12461	2023		47265	Dual-chamber implantable pacemaker, rate-responsive	CT141	Electrodes/Leads and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	28/02/2023	28/02/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	28/02/2023	28/02/2023
					CT145	Cardiovascular devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	28/02/2023	28/02/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	28/02/2023	28/02/2023
					CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	28/02/2023	28/02/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	28/02/2023	28/02/2023
					CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	28/02/2023	28/02/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	28/02/2023	28/02/2023
					CT233	Surgical	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	28/02/2023	28/02/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	28/02/2023	28/02/2023
					CT2406	Implantable	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	28/02/2023	28/02/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	28/02/2023	28/02/2023
					CT2407	Active implantable	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	28/02/2023	28/02/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	28/02/2023	28/02/2023
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	28/02/2023	28/02/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	28/02/2023	28/02/2023

					CT2727	Active implant devices/ancillaries	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	28/02/2023	28/02/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	28/02/2023	28/02/2023
					CT2750	Implantable cardiac electrical stimulators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	28/02/2023	28/02/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	28/02/2023	28/02/2023
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	28/02/2023	28/02/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	28/02/2023	28/02/2023
					CT2985	Active	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	28/02/2023	28/02/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	28/02/2023	28/02/2023
					CT3001	Implantable metal	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	28/02/2023	28/02/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	28/02/2023	28/02/2023
					CT3002	Implantable synthetic polymer	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	28/02/2023	28/02/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	28/02/2023	28/02/2023
					CT3006	Implantable inorganic material	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	28/02/2023	28/02/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	28/02/2023	28/02/2023
					CT314	Electrically powered	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	28/02/2023	28/02/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	28/02/2023	28/02/2023
					CT321	Long-term surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	28/02/2023	28/02/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	28/02/2023	28/02/2023
					CT449	Generators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	28/02/2023	28/02/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	28/02/2023	28/02/2023
					CT981	Single use	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	28/02/2023	28/02/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	28/02/2023	28/02/2023
					CT983	Surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	28/02/2023	28/02/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	28/02/2023	28/02/2023
12462	2023		47265	Dual-chamber implantable pacemaker, rate-responsive	CT141	Electrodes/Leads and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	20/02/2023	20/02/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	20/02/2023	20/02/2023
					CT145	Cardiovascular devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	20/02/2023	20/02/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	20/02/2023	20/02/2023
					CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	20/02/2023	20/02/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	20/02/2023	20/02/2023
					CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	20/02/2023	20/02/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	20/02/2023	20/02/2023
					CT233	Surgical	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	20/02/2023	20/02/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	20/02/2023	20/02/2023
					CT2406	Implantable	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	20/02/2023	20/02/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	20/02/2023	20/02/2023
					CT2407	Active implantable	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	20/02/2023	20/02/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	20/02/2023	20/02/2023
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	20/02/2023	20/02/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	20/02/2023	20/02/2023
					CT2727	Active implant devices/ancillaries	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	20/02/2023	20/02/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	20/02/2023	20/02/2023
					CT2750	Implantable cardiac electrical stimulators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	20/02/2023	20/02/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	20/02/2023	20/02/2023
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	20/02/2023	20/02/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	20/02/2023	20/02/2023
					CT2985	Active	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	20/02/2023	20/02/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	20/02/2023	20/02/2023
					CT3001	Implantable metal	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	20/02/2023	20/02/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	20/02/2023	20/02/2023
					CT3002	Implantable synthetic polymer	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	20/02/2023	20/02/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	20/02/2023	20/02/2023
					CT3006	Implantable inorganic material	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	20/02/2023	20/02/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	20/02/2023	20/02/2023
					CT314	Electrically powered	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	20/02/2023	20/02/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	20/02/2023	20/02/2023
					CT321	Long-term surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	20/02/2023	20/02/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	20/02/2023	20/02/2023
					CT449	Generators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	20/02/2023	20/02/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	20/02/2023	20/02/2023
					CT981	Single use	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	20/02/2023	20/02/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	20/02/2023	20/02/2023
					CT983	Surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	20/02/2023	20/02/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	20/02/2023	20/02/2023
12463	2023	All other reportable incidents	47265	Dual-chamber implantable pacemaker, rate-responsive	CT141	Electrodes/Leads and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	25/02/2023	25/02/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	25/02/2023	25/02/2023
					CT145	Cardiovascular devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	25/02/2023	25/02/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	25/02/2023	25/02/2023
					CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	25/02/2023	25/02/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	25/02/2023	25/02/2023
					CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	25/02/2023	25/02/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	25/02/2023	25/02/2023
					CT233	Surgical	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	25/02/2023	25/02/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	25/02/2023	25/02/2023
					CT2406	Implantable	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	25/02/2023	25/02/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	25/02/2023	25/02/2023
					CT2407	Active implantable	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	25/02/2023	25/02/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	25/02/2023	25/02/2023
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	25/02/2023	25/02/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	25/02/2023	25/02/2023
					CT2727	Active implant devices/ancillaries	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	25/02/2023	25/02/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	25/02/2023	25/02/2023
					CT2750	Implantable cardiac electrical stimulators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	25/02/2023	25/02/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	25/02/2023	25/02/2023
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	25/02/2023	25/02/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	25/02/2023	25/02/2023
					CT2985	Active	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	25/02/2023	25/02/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	25/02/2023	25/02/2023

					CT3001	Implantable metal	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	25/02/2023	25/02/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	25/02/2023	25/02/2023
					CT3002	Implantable synthetic polymer	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	25/02/2023	25/02/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	25/02/2023	25/02/2023
					CT3006	Implantable inorganic material	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	25/02/2023	25/02/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	25/02/2023	25/02/2023
					CT314	Electrically powered	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	25/02/2023	25/02/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	25/02/2023	25/02/2023
					CT321	Long-term surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	25/02/2023	25/02/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	25/02/2023	25/02/2023
					CT449	Generators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	25/02/2023	25/02/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	25/02/2023	25/02/2023
					CT981	Single use	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	25/02/2023	25/02/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	25/02/2023	25/02/2023
					CT983	Surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	25/02/2023	25/02/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	25/02/2023	25/02/2023
12464	2023		47265	Dual-chamber implantable pacemaker, rate-responsive	CT141	Electrodes/Leads and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	25/02/2023	25/02/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	25/02/2023	25/02/2023
					CT145	Cardiovascular devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	25/02/2023	25/02/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	25/02/2023	25/02/2023
					CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	25/02/2023	25/02/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	25/02/2023	25/02/2023
					CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	25/02/2023	25/02/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	25/02/2023	25/02/2023
					CT233	Surgical	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	25/02/2023	25/02/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	25/02/2023	25/02/2023
					CT2406	Implantable	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	25/02/2023	25/02/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	25/02/2023	25/02/2023
					CT2407	Active implantable	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	25/02/2023	25/02/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	25/02/2023	25/02/2023
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	25/02/2023	25/02/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	25/02/2023	25/02/2023
					CT2727	Active implant devices/ancillaries	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	25/02/2023	25/02/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	25/02/2023	25/02/2023
					CT2750	Implantable cardiac electrical stimulators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	25/02/2023	25/02/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	25/02/2023	25/02/2023
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	25/02/2023	25/02/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	25/02/2023	25/02/2023
					CT2985	Active	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	25/02/2023	25/02/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	25/02/2023	25/02/2023
					CT3001	Implantable metal	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	25/02/2023	25/02/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	25/02/2023	25/02/2023
					CT3002	Implantable synthetic polymer	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	25/02/2023	25/02/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	25/02/2023	25/02/2023
					CT3006	Implantable inorganic material	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	25/02/2023	25/02/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	25/02/2023	25/02/2023
					CT314	Electrically powered	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	25/02/2023	25/02/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	25/02/2023	25/02/2023
					CT321	Long-term surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	25/02/2023	25/02/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	25/02/2023	25/02/2023
					CT449	Generators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	25/02/2023	25/02/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	25/02/2023	25/02/2023
					CT981	Single use	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	25/02/2023	25/02/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	25/02/2023	25/02/2023
					CT983	Surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	25/02/2023	25/02/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	25/02/2023	25/02/2023
12465	2023		47265	Dual-chamber implantable pacemaker, rate-responsive	CT141	Electrodes/Leads and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	25/02/2023	25/02/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	25/02/2023	25/02/2023
					CT145	Cardiovascular devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	25/02/2023	25/02/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	25/02/2023	25/02/2023
					CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	25/02/2023	25/02/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	25/02/2023	25/02/2023
					CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	25/02/2023	25/02/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	25/02/2023	25/02/2023
					CT233	Surgical	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	25/02/2023	25/02/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	25/02/2023	25/02/2023
					CT2406	Implantable	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	25/02/2023	25/02/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	25/02/2023	25/02/2023
					CT2407	Active implantable	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	25/02/2023	25/02/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	25/02/2023	25/02/2023
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	25/02/2023	25/02/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	25/02/2023	25/02/2023
					CT2727	Active implant devices/ancillaries	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	25/02/2023	25/02/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	25/02/2023	25/02/2023
					CT2750	Implantable cardiac electrical stimulators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	25/02/2023	25/02/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	25/02/2023	25/02/2023
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	25/02/2023	25/02/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	25/02/2023	25/02/2023
					CT2985	Active	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	25/02/2023	25/02/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	25/02/2023	25/02/2023
					CT3001	Implantable metal	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	25/02/2023	25/02/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	25/02/2023	25/02/2023
					CT3002	Implantable synthetic polymer	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	25/02/2023	25/02/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	25/02/2023	25/02/2023
					CT3006	Implantable inorganic material	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	25/02/2023	25/02/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	25/02/2023	25/02/2023
					CT314	Electrically powered	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	25/02/2023	25/02/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	25/02/2023	25/02/2023

					CT321	Long-term surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	25/02/2023	25/02/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	25/02/2023	25/02/2023
					CT449	Generators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	25/02/2023	25/02/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	25/02/2023	25/02/2023
					CT981	Single use	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	25/02/2023	25/02/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	25/02/2023	25/02/2023
					CT983	Surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	25/02/2023	25/02/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	25/02/2023	25/02/2023
12466	2023		47265	Dual-chamber implantable pacemaker, rate-responsive	CT141	Electrodes/Leads and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	25/02/2023	25/02/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	25/02/2023	25/02/2023
					CT145	Cardiovascular devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	25/02/2023	25/02/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	25/02/2023	25/02/2023
					CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	25/02/2023	25/02/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	25/02/2023	25/02/2023
					CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	25/02/2023	25/02/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	25/02/2023	25/02/2023
					CT233	Surgical	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	25/02/2023	25/02/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	25/02/2023	25/02/2023
					CT2406	Implantable	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	25/02/2023	25/02/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	25/02/2023	25/02/2023
					CT2407	Active implantable	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	25/02/2023	25/02/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	25/02/2023	25/02/2023
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	25/02/2023	25/02/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	25/02/2023	25/02/2023
					CT2727	Active implant devices/ancillaries	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	25/02/2023	25/02/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	25/02/2023	25/02/2023
					CT2750	Implantable cardiac electrical stimulators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	25/02/2023	25/02/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	25/02/2023	25/02/2023
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	25/02/2023	25/02/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	25/02/2023	25/02/2023
					CT2985	Active	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	25/02/2023	25/02/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	25/02/2023	25/02/2023
					CT3001	Implantable metal	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	25/02/2023	25/02/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	25/02/2023	25/02/2023
					CT3002	Implantable synthetic polymer	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	25/02/2023	25/02/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	25/02/2023	25/02/2023
					CT3006	Implantable inorganic material	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	25/02/2023	25/02/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	25/02/2023	25/02/2023
					CT314	Electrically powered	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	25/02/2023	25/02/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	25/02/2023	25/02/2023
					CT321	Long-term surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	25/02/2023	25/02/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	25/02/2023	25/02/2023
					CT449	Generators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	25/02/2023	25/02/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	25/02/2023	25/02/2023
					CT981	Single use	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	25/02/2023	25/02/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	25/02/2023	25/02/2023
					CT983	Surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	25/02/2023	25/02/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	25/02/2023	25/02/2023
12467	2023	All other reportable incidents	47265	Dual-chamber implantable pacemaker, rate-responsive	CT141	Electrodes/Leads and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	13/12/2022	13/12/2022
									A071102	Failure to Interrogate	13/12/2022	13/12/2022
							F1901	Additional Surgery	A070504	Premature Discharge of Battery	13/12/2022	13/12/2022
									A071102	Failure to Interrogate	13/12/2022	13/12/2022
					CT145	Cardiovascular devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	13/12/2022	13/12/2022
									A071102	Failure to Interrogate	13/12/2022	13/12/2022
							F1901	Additional Surgery	A070504	Premature Discharge of Battery	13/12/2022	13/12/2022
									A071102	Failure to Interrogate	13/12/2022	13/12/2022
					CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	13/12/2022	13/12/2022
									A071102	Failure to Interrogate	13/12/2022	13/12/2022
							F1901	Additional Surgery	A070504	Premature Discharge of Battery	13/12/2022	13/12/2022
									A071102	Failure to Interrogate	13/12/2022	13/12/2022
					CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	13/12/2022	13/12/2022
									A071102	Failure to Interrogate	13/12/2022	13/12/2022
							F1901	Additional Surgery	A070504	Premature Discharge of Battery	13/12/2022	13/12/2022
									A071102	Failure to Interrogate	13/12/2022	13/12/2022
					CT233	Surgical	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	13/12/2022	13/12/2022
									A071102	Failure to Interrogate	13/12/2022	13/12/2022
							F1901	Additional Surgery	A070504	Premature Discharge of Battery	13/12/2022	13/12/2022
									A071102	Failure to Interrogate	13/12/2022	13/12/2022
					CT2406	Implantable	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	13/12/2022	13/12/2022
									A071102	Failure to Interrogate	13/12/2022	13/12/2022
							F1901	Additional Surgery	A070504	Premature Discharge of Battery	13/12/2022	13/12/2022
									A071102	Failure to Interrogate	13/12/2022	13/12/2022
					CT2407	Active implantable	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	13/12/2022	13/12/2022
									A071102	Failure to Interrogate	13/12/2022	13/12/2022
							F1901	Additional Surgery	A070504	Premature Discharge of Battery	13/12/2022	13/12/2022
									A071102	Failure to Interrogate	13/12/2022	13/12/2022
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	13/12/2022	13/12/2022
									A071102	Failure to Interrogate	13/12/2022	13/12/2022
							F1901	Additional Surgery	A070504	Premature Discharge of Battery	13/12/2022	13/12/2022
									A071102	Failure to Interrogate	13/12/2022	13/12/2022
					CT2727	Active implant devices/ancillaries	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	13/12/2022	13/12/2022
									A071102	Failure to Interrogate	13/12/2022	13/12/2022
							F1901	Additional Surgery	A070504	Premature Discharge of Battery	13/12/2022	13/12/2022
									A071102	Failure to Interrogate	13/12/2022	13/12/2022
					CT2750	Implantable cardiac electrical stimulators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	13/12/2022	13/12/2022
									A071102	Failure to Interrogate	13/12/2022	13/12/2022
							F1901	Additional Surgery	A070504	Premature Discharge of Battery	13/12/2022	13/12/2022
									A071102	Failure to Interrogate	13/12/2022	13/12/2022

					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	13/12/2022	13/12/2022
									A071102	Failure to Interrogate	13/12/2022	13/12/2022
							F1901	Additional Surgery	A070504	Premature Discharge of Battery	13/12/2022	13/12/2022
									A071102	Failure to Interrogate	13/12/2022	13/12/2022
					CT2985	Active	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	13/12/2022	13/12/2022
									A071102	Failure to Interrogate	13/12/2022	13/12/2022
							F1901	Additional Surgery	A070504	Premature Discharge of Battery	13/12/2022	13/12/2022
									A071102	Failure to Interrogate	13/12/2022	13/12/2022
					CT3001	Implantable metal	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	13/12/2022	13/12/2022
									A071102	Failure to Interrogate	13/12/2022	13/12/2022
							F1901	Additional Surgery	A070504	Premature Discharge of Battery	13/12/2022	13/12/2022
									A071102	Failure to Interrogate	13/12/2022	13/12/2022
					CT3002	Implantable synthetic polymer	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	13/12/2022	13/12/2022
									A071102	Failure to Interrogate	13/12/2022	13/12/2022
							F1901	Additional Surgery	A070504	Premature Discharge of Battery	13/12/2022	13/12/2022
									A071102	Failure to Interrogate	13/12/2022	13/12/2022
					CT3006	Implantable inorganic material	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	13/12/2022	13/12/2022
									A071102	Failure to Interrogate	13/12/2022	13/12/2022
							F1901	Additional Surgery	A070504	Premature Discharge of Battery	13/12/2022	13/12/2022
									A071102	Failure to Interrogate	13/12/2022	13/12/2022
					CT314	Electrically powered	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	13/12/2022	13/12/2022
									A071102	Failure to Interrogate	13/12/2022	13/12/2022
							F1901	Additional Surgery	A070504	Premature Discharge of Battery	13/12/2022	13/12/2022
									A071102	Failure to Interrogate	13/12/2022	13/12/2022
					CT321	Long-term surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	13/12/2022	13/12/2022
									A071102	Failure to Interrogate	13/12/2022	13/12/2022
							F1901	Additional Surgery	A070504	Premature Discharge of Battery	13/12/2022	13/12/2022
									A071102	Failure to Interrogate	13/12/2022	13/12/2022
					CT449	Generators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	13/12/2022	13/12/2022
									A071102	Failure to Interrogate	13/12/2022	13/12/2022
							F1901	Additional Surgery	A070504	Premature Discharge of Battery	13/12/2022	13/12/2022
									A071102	Failure to Interrogate	13/12/2022	13/12/2022
					CT981	Single use	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	13/12/2022	13/12/2022
									A071102	Failure to Interrogate	13/12/2022	13/12/2022
							F1901	Additional Surgery	A070504	Premature Discharge of Battery	13/12/2022	13/12/2022
									A071102	Failure to Interrogate	13/12/2022	13/12/2022
					CT983	Surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	13/12/2022	13/12/2022
									A071102	Failure to Interrogate	13/12/2022	13/12/2022
							F1901	Additional Surgery	A070504	Premature Discharge of Battery	13/12/2022	13/12/2022
									A071102	Failure to Interrogate	13/12/2022	13/12/2022
12468	2023		47265	Dual-chamber implantable pacemaker, rate-responsive	CT141	Electrodes/Leads and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	21/02/2023	21/02/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	21/02/2023	21/02/2023
					CT145	Cardiovascular devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	21/02/2023	21/02/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	21/02/2023	21/02/2023
					CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	21/02/2023	21/02/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	21/02/2023	21/02/2023
					CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	21/02/2023	21/02/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	21/02/2023	21/02/2023
					CT233	Surgical	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	21/02/2023	21/02/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	21/02/2023	21/02/2023
					CT2406	Implantable	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	21/02/2023	21/02/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	21/02/2023	21/02/2023
					CT2407	Active implantable	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	21/02/2023	21/02/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	21/02/2023	21/02/2023
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	21/02/2023	21/02/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	21/02/2023	21/02/2023
					CT2727	Active implant devices/ancillaries	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	21/02/2023	21/02/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	21/02/2023	21/02/2023
					CT2750	Implantable cardiac electrical stimulators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	21/02/2023	21/02/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	21/02/2023	21/02/2023
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	21/02/2023	21/02/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	21/02/2023	21/02/2023
					CT2985	Active	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	21/02/2023	21/02/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	21/02/2023	21/02/2023
					CT3001	Implantable metal	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	21/02/2023	21/02/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	21/02/2023	21/02/2023
					CT3002	Implantable synthetic polymer	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	21/02/2023	21/02/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	21/02/2023	21/02/2023
					CT3006	Implantable inorganic material	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	21/02/2023	21/02/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	21/02/2023	21/02/2023
					CT314	Electrically powered	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	21/02/2023	21/02/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	21/02/2023	21/02/2023
					CT321	Long-term surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	21/02/2023	21/02/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	21/02/2023	21/02/2023
					CT449	Generators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	21/02/2023	21/02/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	21/02/2023	21/02/2023
					CT981	Single use	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	21/02/2023	21/02/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	21/02/2023	21/02/2023
					CT983	Surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	21/02/2023	21/02/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	21/02/2023	21/02/2023
12469	2023		47265	Dual-chamber implantable pacemaker, rate-responsive	CT141	Electrodes/Leads and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	02/02/2023	02/02/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	02/02/2023	02/02/2023
					CT145	Cardiovascular devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	02/02/2023	02/02/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	02/02/2023	02/02/2023
					CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	02/02/2023	02/02/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	02/02/2023	02/02/2023
					CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	02/02/2023	02/02/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	02/02/2023	02/02/2023

					CT233	Surgical	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	02/02/2023	02/02/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	02/02/2023	02/02/2023
					CT2406	Implantable	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	02/02/2023	02/02/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	02/02/2023	02/02/2023
					CT2407	Active implantable	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	02/02/2023	02/02/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	02/02/2023	02/02/2023
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	02/02/2023	02/02/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	02/02/2023	02/02/2023
					CT2727	Active implant devices/ancillaries	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	02/02/2023	02/02/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	02/02/2023	02/02/2023
					CT2750	Implantable cardiac electrical stimulators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	02/02/2023	02/02/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	02/02/2023	02/02/2023
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	02/02/2023	02/02/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	02/02/2023	02/02/2023
					CT2985	Active	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	02/02/2023	02/02/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	02/02/2023	02/02/2023
					CT3001	Implantable metal	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	02/02/2023	02/02/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	02/02/2023	02/02/2023
					CT3002	Implantable synthetic polymer	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	02/02/2023	02/02/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	02/02/2023	02/02/2023
					CT3006	Implantable inorganic material	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	02/02/2023	02/02/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	02/02/2023	02/02/2023
					CT314	Electrically powered	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	02/02/2023	02/02/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	02/02/2023	02/02/2023
					CT321	Long-term surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	02/02/2023	02/02/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	02/02/2023	02/02/2023
					CT449	Generators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	02/02/2023	02/02/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	02/02/2023	02/02/2023
					CT981	Single use	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	02/02/2023	02/02/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	02/02/2023	02/02/2023
					CT983	Surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	02/02/2023	02/02/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	02/02/2023	02/02/2023
12470	2023		47265	Dual-chamber implantable pacemaker, rate-responsive	CT141	Electrodes/Leads and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	02/02/2023	02/02/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	02/02/2023	02/02/2023
					CT145	Cardiovascular devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	02/02/2023	02/02/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	02/02/2023	02/02/2023
					CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	02/02/2023	02/02/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	02/02/2023	02/02/2023
					CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	02/02/2023	02/02/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	02/02/2023	02/02/2023
					CT233	Surgical	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	02/02/2023	02/02/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	02/02/2023	02/02/2023
					CT2406	Implantable	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	02/02/2023	02/02/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	02/02/2023	02/02/2023
					CT2407	Active implantable	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	02/02/2023	02/02/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	02/02/2023	02/02/2023
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	02/02/2023	02/02/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	02/02/2023	02/02/2023
					CT2727	Active implant devices/ancillaries	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	02/02/2023	02/02/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	02/02/2023	02/02/2023
					CT2750	Implantable cardiac electrical stimulators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	02/02/2023	02/02/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	02/02/2023	02/02/2023
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	02/02/2023	02/02/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	02/02/2023	02/02/2023
					CT2985	Active	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	02/02/2023	02/02/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	02/02/2023	02/02/2023
					CT3001	Implantable metal	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	02/02/2023	02/02/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	02/02/2023	02/02/2023
					CT3002	Implantable synthetic polymer	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	02/02/2023	02/02/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	02/02/2023	02/02/2023
					CT3006	Implantable inorganic material	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	02/02/2023	02/02/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	02/02/2023	02/02/2023
					CT314	Electrically powered	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	02/02/2023	02/02/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	02/02/2023	02/02/2023
					CT321	Long-term surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	02/02/2023	02/02/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	02/02/2023	02/02/2023
					CT449	Generators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	02/02/2023	02/02/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	02/02/2023	02/02/2023
					CT981	Single use	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	02/02/2023	02/02/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	02/02/2023	02/02/2023
					CT983	Surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	02/02/2023	02/02/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	02/02/2023	02/02/2023
12471	2023	All other reportable incidents	47265	Dual-chamber implantable pacemaker, rate-responsive	CT141	Electrodes/Leads and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	13/02/2023	13/02/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	13/02/2023	13/02/2023
					CT145	Cardiovascular devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	13/02/2023	13/02/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	13/02/2023	13/02/2023
					CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	13/02/2023	13/02/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	13/02/2023	13/02/2023
					CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	13/02/2023	13/02/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	13/02/2023	13/02/2023
					CT233	Surgical	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	13/02/2023	13/02/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	13/02/2023	13/02/2023
					CT2406	Implantable	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	13/02/2023	13/02/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	13/02/2023	13/02/2023
					CT2407	Active implantable	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	13/02/2023	13/02/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	13/02/2023	13/02/2023
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	13/02/2023	13/02/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	13/02/2023	13/02/2023

					CT2727	Active implant devices/ancillaries	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	13/02/2023	13/02/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	13/02/2023	13/02/2023
					CT2750	Implantable cardiac electrical stimulators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	13/02/2023	13/02/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	13/02/2023	13/02/2023
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	13/02/2023	13/02/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	13/02/2023	13/02/2023
					CT2985	Active	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	13/02/2023	13/02/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	13/02/2023	13/02/2023
					CT3001	Implantable metal	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	13/02/2023	13/02/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	13/02/2023	13/02/2023
					CT3002	Implantable synthetic polymer	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	13/02/2023	13/02/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	13/02/2023	13/02/2023
					CT3006	Implantable inorganic material	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	13/02/2023	13/02/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	13/02/2023	13/02/2023
					CT314	Electrically powered	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	13/02/2023	13/02/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	13/02/2023	13/02/2023
					CT321	Long-term surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	13/02/2023	13/02/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	13/02/2023	13/02/2023
					CT449	Generators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	13/02/2023	13/02/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	13/02/2023	13/02/2023
					CT981	Single use	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	13/02/2023	13/02/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	13/02/2023	13/02/2023
					CT983	Surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	13/02/2023	13/02/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	13/02/2023	13/02/2023
12472	2023	All other reportable incidents	17882	Physiologic-monitoring defibrillation system	CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070803	Failure to Power Up		
							F27	Problem Identified During Non-Clinical Procedure	A070803	Failure to Power Up		
12473	2023	All other reportable incidents	35223	Endocardial/interventricular septal pacing lead	CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A072201	High Impedance	24/02/2023	24/02/2023
							F08	Hospitalization or Prolonged Hospitalization	A072201	High Impedance	24/02/2023	24/02/2023
							F12	Serious Injury/ Illness/ Impairment	A072201	High Impedance	24/02/2023	24/02/2023
							F1903	Device Explantation	A072201	High Impedance	24/02/2023	24/02/2023
12474	2023	All other reportable incidents	60789	Intracardiac pacemaker	CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A05	Mechanical Problem		
									A0504	Leak/Splash		
							F26	No Health Consequences or Impact	A05	Mechanical Problem		
									A0504	Leak/Splash		
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A05	Mechanical Problem		
									A0504	Leak/Splash		
							F26	No Health Consequences or Impact	A05	Mechanical Problem		
									A0504	Leak/Splash		
					CT2818	Catheterization/Cannulation/Fluid path devices	E2403	No Clinical Signs, Symptoms or Conditions	A05	Mechanical Problem		
									A0504	Leak/Splash		
							F26	No Health Consequences or Impact	A05	Mechanical Problem		
									A0504	Leak/Splash		
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A05	Mechanical Problem		
									A0504	Leak/Splash		
							F26	No Health Consequences or Impact	A05	Mechanical Problem		
									A0504	Leak/Splash		
12475	2023	All other reportable incidents	60789	Intracardiac pacemaker	CT130	Catheters and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem		
							F26	No Health Consequences or Impact	A0705	Battery Problem		
					CT145	Cardiovascular devices	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem		
							F26	No Health Consequences or Impact	A0705	Battery Problem		
					CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem		
							F26	No Health Consequences or Impact	A0705	Battery Problem		
					CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem		
							F26	No Health Consequences or Impact	A0705	Battery Problem		
					CT233	Surgical	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem		
							F26	No Health Consequences or Impact	A0705	Battery Problem		
					CT2406	Implantable	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem		
							F26	No Health Consequences or Impact	A0705	Battery Problem		
					CT2407	Active implantable	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem		
							F26	No Health Consequences or Impact	A0705	Battery Problem		
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem		
							F26	No Health Consequences or Impact	A0705	Battery Problem		
					CT2727	Active implant devices/ancillaries	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem		
							F26	No Health Consequences or Impact	A0705	Battery Problem		
					CT2750	Implantable cardiac electrical stimulators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem		
							F26	No Health Consequences or Impact	A0705	Battery Problem		
					CT2818	Catheterization/Cannulation/Fluid path devices	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem		
							F26	No Health Consequences or Impact	A0705	Battery Problem		
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem		
							F26	No Health Consequences or Impact	A0705	Battery Problem		
					CT2985	Active	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem		
							F26	No Health Consequences or Impact	A0705	Battery Problem		
					CT3001	Implantable metal	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem		
							F26	No Health Consequences or Impact	A0705	Battery Problem		
					CT3002	Implantable synthetic polymer	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem		
							F26	No Health Consequences or Impact	A0705	Battery Problem		
					CT3006	Implantable inorganic material	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem		
							F26	No Health Consequences or Impact	A0705	Battery Problem		
					CT314	Electrically powered	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem		
							F26	No Health Consequences or Impact	A0705	Battery Problem		
					CT321	Long-term surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem		
							F26	No Health Consequences or Impact	A0705	Battery Problem		
					CT981	Single use	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem		
							F26	No Health Consequences or Impact	A0705	Battery Problem		
					CT983	Surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem		
							F26	No Health Consequences or Impact	A0705	Battery Problem		
12476	2023		47265	Dual-chamber implantable pacemaker, rate-responsive	CT141	Electrodes/Leads and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	17/02/2023	17/02/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	17/02/2023	17/02/2023

					CT145	Cardiovascular devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	17/02/2023	17/02/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	17/02/2023	17/02/2023
					CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	17/02/2023	17/02/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	17/02/2023	17/02/2023
					CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	17/02/2023	17/02/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	17/02/2023	17/02/2023
					CT233	Surgical	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	17/02/2023	17/02/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	17/02/2023	17/02/2023
					CT2406	Implantable	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	17/02/2023	17/02/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	17/02/2023	17/02/2023
					CT2407	Active implantable	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	17/02/2023	17/02/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	17/02/2023	17/02/2023
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	17/02/2023	17/02/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	17/02/2023	17/02/2023
					CT2727	Active implant devices/ancillaries	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	17/02/2023	17/02/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	17/02/2023	17/02/2023
					CT2750	Implantable cardiac electrical stimulators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	17/02/2023	17/02/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	17/02/2023	17/02/2023
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	17/02/2023	17/02/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	17/02/2023	17/02/2023
					CT2985	Active	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	17/02/2023	17/02/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	17/02/2023	17/02/2023
					CT3001	Implantable metal	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	17/02/2023	17/02/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	17/02/2023	17/02/2023
					CT3002	Implantable synthetic polymer	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	17/02/2023	17/02/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	17/02/2023	17/02/2023
					CT3006	Implantable inorganic material	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	17/02/2023	17/02/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	17/02/2023	17/02/2023
					CT314	Electrically powered	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	17/02/2023	17/02/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	17/02/2023	17/02/2023
					CT321	Long-term surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	17/02/2023	17/02/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	17/02/2023	17/02/2023
					CT449	Generators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	17/02/2023	17/02/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	17/02/2023	17/02/2023
					CT981	Single use	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	17/02/2023	17/02/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	17/02/2023	17/02/2023
					CT983	Surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	17/02/2023	17/02/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	17/02/2023	17/02/2023
12477	2023		47265	Dual-chamber implantable pacemaker, rate-responsive	CT141	Electrodes/Leads and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	19/02/2023	19/02/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	19/02/2023	19/02/2023
					CT145	Cardiovascular devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	19/02/2023	19/02/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	19/02/2023	19/02/2023
					CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	19/02/2023	19/02/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	19/02/2023	19/02/2023
					CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	19/02/2023	19/02/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	19/02/2023	19/02/2023
					CT233	Surgical	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	19/02/2023	19/02/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	19/02/2023	19/02/2023
					CT2406	Implantable	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	19/02/2023	19/02/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	19/02/2023	19/02/2023
					CT2407	Active implantable	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	19/02/2023	19/02/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	19/02/2023	19/02/2023
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	19/02/2023	19/02/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	19/02/2023	19/02/2023
					CT2727	Active implant devices/ancillaries	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	19/02/2023	19/02/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	19/02/2023	19/02/2023
					CT2750	Implantable cardiac electrical stimulators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	19/02/2023	19/02/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	19/02/2023	19/02/2023
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	19/02/2023	19/02/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	19/02/2023	19/02/2023
					CT2985	Active	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	19/02/2023	19/02/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	19/02/2023	19/02/2023
					CT3001	Implantable metal	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	19/02/2023	19/02/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	19/02/2023	19/02/2023
					CT3002	Implantable synthetic polymer	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	19/02/2023	19/02/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	19/02/2023	19/02/2023
					CT3006	Implantable inorganic material	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	19/02/2023	19/02/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	19/02/2023	19/02/2023
					CT314	Electrically powered	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	19/02/2023	19/02/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	19/02/2023	19/02/2023
					CT321	Long-term surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	19/02/2023	19/02/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	19/02/2023	19/02/2023
					CT449	Generators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	19/02/2023	19/02/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	19/02/2023	19/02/2023
					CT981	Single use	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	19/02/2023	19/02/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	19/02/2023	19/02/2023
					CT983	Surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	19/02/2023	19/02/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	19/02/2023	19/02/2023
12478	2023	All other reportable incidents	35852	Single-chamber implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
									A0908	Incorrect, Inadequate or Imprecise Result or Readings		
							F26	No Health Consequences or Impact	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
									A0908	Incorrect, Inadequate or Imprecise Result or Readings		
12479	2023		47265	Dual-chamber implantable pacemaker, rate-responsive	CT141	Electrodes/Leads and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	19/02/2023	19/02/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	19/02/2023	19/02/2023

					CT145	Cardiovascular devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	19/02/2023	19/02/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	19/02/2023	19/02/2023
					CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	19/02/2023	19/02/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	19/02/2023	19/02/2023
					CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	19/02/2023	19/02/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	19/02/2023	19/02/2023
					CT233	Surgical	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	19/02/2023	19/02/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	19/02/2023	19/02/2023
					CT2406	Implantable	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	19/02/2023	19/02/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	19/02/2023	19/02/2023
					CT2407	Active implantable	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	19/02/2023	19/02/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	19/02/2023	19/02/2023
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	19/02/2023	19/02/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	19/02/2023	19/02/2023
					CT2727	Active implant devices/ancillaries	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	19/02/2023	19/02/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	19/02/2023	19/02/2023
					CT2750	Implantable cardiac electrical stimulators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	19/02/2023	19/02/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	19/02/2023	19/02/2023
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	19/02/2023	19/02/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	19/02/2023	19/02/2023
					CT2985	Active	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	19/02/2023	19/02/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	19/02/2023	19/02/2023
					CT3001	Implantable metal	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	19/02/2023	19/02/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	19/02/2023	19/02/2023
					CT3002	Implantable synthetic polymer	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	19/02/2023	19/02/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	19/02/2023	19/02/2023
					CT3006	Implantable inorganic material	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	19/02/2023	19/02/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	19/02/2023	19/02/2023
					CT314	Electrically powered	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	19/02/2023	19/02/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	19/02/2023	19/02/2023
					CT321	Long-term surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	19/02/2023	19/02/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	19/02/2023	19/02/2023
					CT449	Generators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	19/02/2023	19/02/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	19/02/2023	19/02/2023
					CT981	Single use	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	19/02/2023	19/02/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	19/02/2023	19/02/2023
					CT983	Surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	19/02/2023	19/02/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	19/02/2023	19/02/2023
12480	2023	All other reportable incidents	17882	Physiologic-monitoring defibrillation system	CT1684	Cardiac pacemakers and associated devices	E2401	Insufficient Information	A110208	Unintended Application Program Shut Down		
							F26	No Health Consequences or Impact	A110208	Unintended Application Program Shut Down		
12481	2023		47265	Dual-chamber implantable pacemaker, rate-responsive	CT141	Electrodes/Leads and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A25	No Apparent Adverse Event		
							F26	No Health Consequences or Impact	A25	No Apparent Adverse Event		
					CT145	Cardiovascular devices	E2403	No Clinical Signs, Symptoms or Conditions	A25	No Apparent Adverse Event		
							F26	No Health Consequences or Impact	A25	No Apparent Adverse Event		
					CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A25	No Apparent Adverse Event		
							F26	No Health Consequences or Impact	A25	No Apparent Adverse Event		
					CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A25	No Apparent Adverse Event		
							F26	No Health Consequences or Impact	A25	No Apparent Adverse Event		
					CT233	Surgical	E2403	No Clinical Signs, Symptoms or Conditions	A25	No Apparent Adverse Event		
							F26	No Health Consequences or Impact	A25	No Apparent Adverse Event		
					CT2406	Implantable	E2403	No Clinical Signs, Symptoms or Conditions	A25	No Apparent Adverse Event		
							F26	No Health Consequences or Impact	A25	No Apparent Adverse Event		
					CT2407	Active implantable	E2403	No Clinical Signs, Symptoms or Conditions	A25	No Apparent Adverse Event		
							F26	No Health Consequences or Impact	A25	No Apparent Adverse Event		
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A25	No Apparent Adverse Event		
							F26	No Health Consequences or Impact	A25	No Apparent Adverse Event		
					CT2727	Active implant devices/ancillaries	E2403	No Clinical Signs, Symptoms or Conditions	A25	No Apparent Adverse Event		
							F26	No Health Consequences or Impact	A25	No Apparent Adverse Event		
					CT2750	Implantable cardiac electrical stimulators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A25	No Apparent Adverse Event		
							F26	No Health Consequences or Impact	A25	No Apparent Adverse Event		
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A25	No Apparent Adverse Event		
							F26	No Health Consequences or Impact	A25	No Apparent Adverse Event		
					CT2985	Active	E2403	No Clinical Signs, Symptoms or Conditions	A25	No Apparent Adverse Event		
							F26	No Health Consequences or Impact	A25	No Apparent Adverse Event		
					CT3001	Implantable metal	E2403	No Clinical Signs, Symptoms or Conditions	A25	No Apparent Adverse Event		
							F26	No Health Consequences or Impact	A25	No Apparent Adverse Event		
					CT3002	Implantable synthetic polymer	E2403	No Clinical Signs, Symptoms or Conditions	A25	No Apparent Adverse Event		
							F26	No Health Consequences or Impact	A25	No Apparent Adverse Event		
					CT3006	Implantable inorganic material	E2403	No Clinical Signs, Symptoms or Conditions	A25	No Apparent Adverse Event		
							F26	No Health Consequences or Impact	A25	No Apparent Adverse Event		
					CT314	Electrically powered	E2403	No Clinical Signs, Symptoms or Conditions	A25	No Apparent Adverse Event		
							F26	No Health Consequences or Impact	A25	No Apparent Adverse Event		
					CT321	Long-term surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A25	No Apparent Adverse Event		
							F26	No Health Consequences or Impact	A25	No Apparent Adverse Event		
					CT449	Generators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A25	No Apparent Adverse Event		
							F26	No Health Consequences or Impact	A25	No Apparent Adverse Event		
					CT981	Single use	E2403	No Clinical Signs, Symptoms or Conditions	A25	No Apparent Adverse Event		
							F26	No Health Consequences or Impact	A25	No Apparent Adverse Event		
					CT983	Surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A25	No Apparent Adverse Event		
							F26	No Health Consequences or Impact	A25	No Apparent Adverse Event		
12482	2023		47265	Dual-chamber implantable pacemaker, rate-responsive	CT141	Electrodes/Leads and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	04/02/2023	04/02/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	04/02/2023	04/02/2023
					CT145	Cardiovascular devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	04/02/2023	04/02/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	04/02/2023	04/02/2023
					CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	04/02/2023	04/02/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	04/02/2023	04/02/2023
					CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	04/02/2023	04/02/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	04/02/2023	04/02/2023

					CT233	Surgical	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	04/02/2023	04/02/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	04/02/2023	04/02/2023
					CT2406	Implantable	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	04/02/2023	04/02/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	04/02/2023	04/02/2023
					CT2407	Active implantable	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	04/02/2023	04/02/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	04/02/2023	04/02/2023
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	04/02/2023	04/02/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	04/02/2023	04/02/2023
					CT2727	Active implant devices/ancillaries	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	04/02/2023	04/02/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	04/02/2023	04/02/2023
					CT2750	Implantable cardiac electrical stimulators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	04/02/2023	04/02/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	04/02/2023	04/02/2023
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	04/02/2023	04/02/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	04/02/2023	04/02/2023
					CT2985	Active	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	04/02/2023	04/02/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	04/02/2023	04/02/2023
					CT3001	Implantable metal	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	04/02/2023	04/02/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	04/02/2023	04/02/2023
					CT3002	Implantable synthetic polymer	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	04/02/2023	04/02/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	04/02/2023	04/02/2023
					CT3006	Implantable inorganic material	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	04/02/2023	04/02/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	04/02/2023	04/02/2023
					CT314	Electrically powered	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	04/02/2023	04/02/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	04/02/2023	04/02/2023
					CT321	Long-term surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	04/02/2023	04/02/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	04/02/2023	04/02/2023
					CT449	Generators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	04/02/2023	04/02/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	04/02/2023	04/02/2023
					CT981	Single use	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	04/02/2023	04/02/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	04/02/2023	04/02/2023
					CT983	Surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	04/02/2023	04/02/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	04/02/2023	04/02/2023
12483	2023		47265	Dual-chamber implantable pacemaker, rate-responsive	CT141	Electrodes/Leads and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	15/02/2023	15/02/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	15/02/2023	15/02/2023
					CT145	Cardiovascular devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	15/02/2023	15/02/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	15/02/2023	15/02/2023
					CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	15/02/2023	15/02/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	15/02/2023	15/02/2023
					CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	15/02/2023	15/02/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	15/02/2023	15/02/2023
					CT233	Surgical	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	15/02/2023	15/02/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	15/02/2023	15/02/2023
					CT2406	Implantable	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	15/02/2023	15/02/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	15/02/2023	15/02/2023
					CT2407	Active implantable	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	15/02/2023	15/02/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	15/02/2023	15/02/2023
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	15/02/2023	15/02/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	15/02/2023	15/02/2023
					CT2727	Active implant devices/ancillaries	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	15/02/2023	15/02/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	15/02/2023	15/02/2023
					CT2750	Implantable cardiac electrical stimulators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	15/02/2023	15/02/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	15/02/2023	15/02/2023
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	15/02/2023	15/02/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	15/02/2023	15/02/2023
					CT2985	Active	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	15/02/2023	15/02/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	15/02/2023	15/02/2023
					CT3001	Implantable metal	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	15/02/2023	15/02/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	15/02/2023	15/02/2023
					CT3002	Implantable synthetic polymer	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	15/02/2023	15/02/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	15/02/2023	15/02/2023
					CT3006	Implantable inorganic material	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	15/02/2023	15/02/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	15/02/2023	15/02/2023
					CT314	Electrically powered	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	15/02/2023	15/02/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	15/02/2023	15/02/2023
					CT321	Long-term surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	15/02/2023	15/02/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	15/02/2023	15/02/2023
					CT449	Generators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	15/02/2023	15/02/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	15/02/2023	15/02/2023
					CT981	Single use	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	15/02/2023	15/02/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	15/02/2023	15/02/2023
					CT983	Surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	15/02/2023	15/02/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	15/02/2023	15/02/2023
12484	2023		47265	Dual-chamber implantable pacemaker, rate-responsive	CT141	Electrodes/Leads and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	24/02/2023	24/02/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	24/02/2023	24/02/2023
					CT145	Cardiovascular devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	24/02/2023	24/02/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	24/02/2023	24/02/2023
					CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	24/02/2023	24/02/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	24/02/2023	24/02/2023
					CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	24/02/2023	24/02/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	24/02/2023	24/02/2023
					CT233	Surgical	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	24/02/2023	24/02/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	24/02/2023	24/02/2023
					CT2406	Implantable	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	24/02/2023	24/02/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	24/02/2023	24/02/2023
					CT2407	Active implantable	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	24/02/2023	24/02/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	24/02/2023	24/02/2023
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	24/02/2023	24/02/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	24/02/2023	24/02/2023

					CT2727	Active implant devices/ancillaries	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	24/02/2023	24/02/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	24/02/2023	24/02/2023
					CT2750	Implantable cardiac electrical stimulators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	24/02/2023	24/02/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	24/02/2023	24/02/2023
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	24/02/2023	24/02/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	24/02/2023	24/02/2023
					CT2985	Active	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	24/02/2023	24/02/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	24/02/2023	24/02/2023
					CT3001	Implantable metal	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	24/02/2023	24/02/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	24/02/2023	24/02/2023
					CT3002	Implantable synthetic polymer	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	24/02/2023	24/02/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	24/02/2023	24/02/2023
					CT3006	Implantable inorganic material	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	24/02/2023	24/02/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	24/02/2023	24/02/2023
					CT314	Electrically powered	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	24/02/2023	24/02/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	24/02/2023	24/02/2023
					CT321	Long-term surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	24/02/2023	24/02/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	24/02/2023	24/02/2023
					CT449	Generators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	24/02/2023	24/02/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	24/02/2023	24/02/2023
					CT981	Single use	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	24/02/2023	24/02/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	24/02/2023	24/02/2023
					CT983	Surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	24/02/2023	24/02/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	24/02/2023	24/02/2023
12485	2023		47265	Dual-chamber implantable pacemaker, rate-responsive	CT141	Electrodes/Leads and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	23/02/2023	23/02/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	23/02/2023	23/02/2023
					CT145	Cardiovascular devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	23/02/2023	23/02/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	23/02/2023	23/02/2023
					CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	23/02/2023	23/02/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	23/02/2023	23/02/2023
					CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	23/02/2023	23/02/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	23/02/2023	23/02/2023
					CT233	Surgical	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	23/02/2023	23/02/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	23/02/2023	23/02/2023
					CT2406	Implantable	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	23/02/2023	23/02/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	23/02/2023	23/02/2023
					CT2407	Active implantable	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	23/02/2023	23/02/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	23/02/2023	23/02/2023
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	23/02/2023	23/02/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	23/02/2023	23/02/2023
					CT2727	Active implant devices/ancillaries	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	23/02/2023	23/02/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	23/02/2023	23/02/2023
					CT2750	Implantable cardiac electrical stimulators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	23/02/2023	23/02/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	23/02/2023	23/02/2023
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	23/02/2023	23/02/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	23/02/2023	23/02/2023
					CT2985	Active	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	23/02/2023	23/02/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	23/02/2023	23/02/2023
					CT3001	Implantable metal	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	23/02/2023	23/02/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	23/02/2023	23/02/2023
					CT3002	Implantable synthetic polymer	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	23/02/2023	23/02/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	23/02/2023	23/02/2023
					CT3006	Implantable inorganic material	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	23/02/2023	23/02/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	23/02/2023	23/02/2023
					CT314	Electrically powered	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	23/02/2023	23/02/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	23/02/2023	23/02/2023
					CT321	Long-term surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	23/02/2023	23/02/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	23/02/2023	23/02/2023
					CT449	Generators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	23/02/2023	23/02/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	23/02/2023	23/02/2023
					CT981	Single use	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	23/02/2023	23/02/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	23/02/2023	23/02/2023
					CT983	Surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	23/02/2023	23/02/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	23/02/2023	23/02/2023
12486	2023		47265	Dual-chamber implantable pacemaker, rate-responsive	CT141	Electrodes/Leads and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	23/02/2023	23/02/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	23/02/2023	23/02/2023
					CT145	Cardiovascular devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	23/02/2023	23/02/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	23/02/2023	23/02/2023
					CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	23/02/2023	23/02/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	23/02/2023	23/02/2023
					CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	23/02/2023	23/02/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	23/02/2023	23/02/2023
					CT233	Surgical	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	23/02/2023	23/02/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	23/02/2023	23/02/2023
					CT2406	Implantable	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	23/02/2023	23/02/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	23/02/2023	23/02/2023
					CT2407	Active implantable	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	23/02/2023	23/02/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	23/02/2023	23/02/2023
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	23/02/2023	23/02/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	23/02/2023	23/02/2023
					CT2727	Active implant devices/ancillaries	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	23/02/2023	23/02/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	23/02/2023	23/02/2023
					CT2750	Implantable cardiac electrical stimulators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	23/02/2023	23/02/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	23/02/2023	23/02/2023
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	23/02/2023	23/02/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	23/02/2023	23/02/2023
					CT2985	Active	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	23/02/2023	23/02/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	23/02/2023	23/02/2023

					CT3001	Implantable metal	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	23/02/2023	23/02/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	23/02/2023	23/02/2023
					CT3002	Implantable synthetic polymer	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	23/02/2023	23/02/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	23/02/2023	23/02/2023
					CT3006	Implantable inorganic material	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	23/02/2023	23/02/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	23/02/2023	23/02/2023
					CT314	Electrically powered	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	23/02/2023	23/02/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	23/02/2023	23/02/2023
					CT321	Long-term surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	23/02/2023	23/02/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	23/02/2023	23/02/2023
					CT449	Generators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	23/02/2023	23/02/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	23/02/2023	23/02/2023
					CT981	Single use	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	23/02/2023	23/02/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	23/02/2023	23/02/2023
					CT983	Surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	23/02/2023	23/02/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	23/02/2023	23/02/2023
12487	2023		47265	Dual-chamber implantable pacemaker, rate-responsive	CT141	Electrodes/Leads and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	28/02/2023	28/02/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	28/02/2023	28/02/2023
					CT145	Cardiovascular devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	28/02/2023	28/02/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	28/02/2023	28/02/2023
					CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	28/02/2023	28/02/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	28/02/2023	28/02/2023
					CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	28/02/2023	28/02/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	28/02/2023	28/02/2023
					CT233	Surgical	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	28/02/2023	28/02/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	28/02/2023	28/02/2023
					CT2406	Implantable	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	28/02/2023	28/02/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	28/02/2023	28/02/2023
					CT2407	Active implantable	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	28/02/2023	28/02/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	28/02/2023	28/02/2023
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	28/02/2023	28/02/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	28/02/2023	28/02/2023
					CT2727	Active implant devices/ancillaries	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	28/02/2023	28/02/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	28/02/2023	28/02/2023
					CT2750	Implantable cardiac electrical stimulators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	28/02/2023	28/02/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	28/02/2023	28/02/2023
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	28/02/2023	28/02/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	28/02/2023	28/02/2023
					CT2985	Active	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	28/02/2023	28/02/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	28/02/2023	28/02/2023
					CT3001	Implantable metal	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	28/02/2023	28/02/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	28/02/2023	28/02/2023
					CT3002	Implantable synthetic polymer	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	28/02/2023	28/02/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	28/02/2023	28/02/2023
					CT3006	Implantable inorganic material	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	28/02/2023	28/02/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	28/02/2023	28/02/2023
					CT314	Electrically powered	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	28/02/2023	28/02/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	28/02/2023	28/02/2023
					CT321	Long-term surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	28/02/2023	28/02/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	28/02/2023	28/02/2023
					CT449	Generators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	28/02/2023	28/02/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	28/02/2023	28/02/2023
					CT981	Single use	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	28/02/2023	28/02/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	28/02/2023	28/02/2023
					CT983	Surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	28/02/2023	28/02/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	28/02/2023	28/02/2023
12488	2023	All other reportable incidents	35852	Single-chamber implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E2104	Electric Shock	A070909	Oversensing		
									A071302	Inappropriate/Inadequate Shock/Stimulation		
									A090801	Signal Artifact/Noise		
									A1301	Failure to Read Input Signal		
							F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A070909	Oversensing		
									A071302	Inappropriate/Inadequate Shock/Stimulation		
									A090801	Signal Artifact/Noise		
									A1301	Failure to Read Input Signal		
							F15	Recognized Device or Procedural Complication	A070909	Oversensing		
									A071302	Inappropriate/Inadequate Shock/Stimulation		
									A090801	Signal Artifact/Noise		
									A1301	Failure to Read Input Signal		
							F2203	Imaging Required	A070909	Oversensing		
									A071302	Inappropriate/Inadequate Shock/Stimulation		
									A090801	Signal Artifact/Noise		
									A1301	Failure to Read Input Signal		
12489	2023		47265	Dual-chamber implantable pacemaker, rate-responsive	CT141	Electrodes/Leads and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	25/02/2023	25/02/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	25/02/2023	25/02/2023
					CT145	Cardiovascular devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	25/02/2023	25/02/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	25/02/2023	25/02/2023
					CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	25/02/2023	25/02/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	25/02/2023	25/02/2023
					CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	25/02/2023	25/02/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	25/02/2023	25/02/2023
					CT233	Surgical	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	25/02/2023	25/02/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	25/02/2023	25/02/2023
					CT2406	Implantable	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	25/02/2023	25/02/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	25/02/2023	25/02/2023
					CT2407	Active implantable	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	25/02/2023	25/02/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	25/02/2023	25/02/2023
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	25/02/2023	25/02/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	25/02/2023	25/02/2023

					CT2727	Active implant devices/ancillaries	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	25/02/2023	25/02/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	25/02/2023	25/02/2023
					CT2750	Implantable cardiac electrical stimulators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	25/02/2023	25/02/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	25/02/2023	25/02/2023
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	25/02/2023	25/02/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	25/02/2023	25/02/2023
					CT2985	Active	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	25/02/2023	25/02/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	25/02/2023	25/02/2023
					CT3001	Implantable metal	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	25/02/2023	25/02/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	25/02/2023	25/02/2023
					CT3002	Implantable synthetic polymer	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	25/02/2023	25/02/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	25/02/2023	25/02/2023
					CT3006	Implantable inorganic material	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	25/02/2023	25/02/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	25/02/2023	25/02/2023
					CT314	Electrically powered	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	25/02/2023	25/02/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	25/02/2023	25/02/2023
					CT321	Long-term surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	25/02/2023	25/02/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	25/02/2023	25/02/2023
					CT449	Generators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	25/02/2023	25/02/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	25/02/2023	25/02/2023
					CT981	Single use	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	25/02/2023	25/02/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	25/02/2023	25/02/2023
					CT983	Surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	25/02/2023	25/02/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	25/02/2023	25/02/2023
12490	2023	All other reportable incidents	35852	Single-chamber implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E2104	Electric Shock	A070909	Oversensing		
									A071302	Inappropriate/Inadequate Shock/Stimulation		
									A090801	Signal Artifact/Noise		
									A1301	Failure to Read Input Signal		
							F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A070909	Oversensing		
									A071302	Inappropriate/Inadequate Shock/Stimulation		
									A090801	Signal Artifact/Noise		
									A1301	Failure to Read Input Signal		
							F15	Recognized Device or Procedural Complication	A070909	Oversensing		
									A071302	Inappropriate/Inadequate Shock/Stimulation		
									A090801	Signal Artifact/Noise		
									A1301	Failure to Read Input Signal		
							F2203	Imaging Required	A070909	Oversensing		
									A071302	Inappropriate/Inadequate Shock/Stimulation		
									A090801	Signal Artifact/Noise		
									A1301	Failure to Read Input Signal		
12491	2023		47265	Dual-chamber implantable pacemaker, rate-responsive	CT141	Electrodes/Leads and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	25/02/2023	25/02/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	25/02/2023	25/02/2023
					CT145	Cardiovascular devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	25/02/2023	25/02/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	25/02/2023	25/02/2023
					CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	25/02/2023	25/02/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	25/02/2023	25/02/2023
					CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	25/02/2023	25/02/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	25/02/2023	25/02/2023
					CT233	Surgical	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	25/02/2023	25/02/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	25/02/2023	25/02/2023
					CT2406	Implantable	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	25/02/2023	25/02/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	25/02/2023	25/02/2023
					CT2407	Active implantable	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	25/02/2023	25/02/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	25/02/2023	25/02/2023
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	25/02/2023	25/02/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	25/02/2023	25/02/2023
					CT2727	Active implant devices/ancillaries	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	25/02/2023	25/02/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	25/02/2023	25/02/2023
					CT2750	Implantable cardiac electrical stimulators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	25/02/2023	25/02/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	25/02/2023	25/02/2023
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	25/02/2023	25/02/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	25/02/2023	25/02/2023
					CT2985	Active	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	25/02/2023	25/02/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	25/02/2023	25/02/2023
					CT3001	Implantable metal	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	25/02/2023	25/02/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	25/02/2023	25/02/2023
					CT3002	Implantable synthetic polymer	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	25/02/2023	25/02/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	25/02/2023	25/02/2023
					CT3006	Implantable inorganic material	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	25/02/2023	25/02/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	25/02/2023	25/02/2023
					CT314	Electrically powered	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	25/02/2023	25/02/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	25/02/2023	25/02/2023
					CT321	Long-term surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	25/02/2023	25/02/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	25/02/2023	25/02/2023
					CT449	Generators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	25/02/2023	25/02/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	25/02/2023	25/02/2023
					CT981	Single use	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	25/02/2023	25/02/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	25/02/2023	25/02/2023
					CT983	Surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	25/02/2023	25/02/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	25/02/2023	25/02/2023
12492	2023		47265	Dual-chamber implantable pacemaker, rate-responsive	CT141	Electrodes/Leads and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	25/02/2023	25/02/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	25/02/2023	25/02/2023
					CT145	Cardiovascular devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	25/02/2023	25/02/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	25/02/2023	25/02/2023
					CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	25/02/2023	25/02/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	25/02/2023	25/02/2023
					CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	25/02/2023	25/02/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	25/02/2023	25/02/2023

					CT233	Surgical	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	25/02/2023	25/02/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	25/02/2023	25/02/2023
					CT2406	Implantable	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	25/02/2023	25/02/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	25/02/2023	25/02/2023
					CT2407	Active implantable	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	25/02/2023	25/02/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	25/02/2023	25/02/2023
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	25/02/2023	25/02/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	25/02/2023	25/02/2023
					CT2727	Active implant devices/ancillaries	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	25/02/2023	25/02/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	25/02/2023	25/02/2023
					CT2750	Implantable cardiac electrical stimulators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	25/02/2023	25/02/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	25/02/2023	25/02/2023
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	25/02/2023	25/02/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	25/02/2023	25/02/2023
					CT2985	Active	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	25/02/2023	25/02/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	25/02/2023	25/02/2023
					CT3001	Implantable metal	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	25/02/2023	25/02/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	25/02/2023	25/02/2023
					CT3002	Implantable synthetic polymer	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	25/02/2023	25/02/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	25/02/2023	25/02/2023
					CT3006	Implantable inorganic material	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	25/02/2023	25/02/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	25/02/2023	25/02/2023
					CT314	Electrically powered	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	25/02/2023	25/02/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	25/02/2023	25/02/2023
					CT321	Long-term surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	25/02/2023	25/02/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	25/02/2023	25/02/2023
					CT449	Generators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	25/02/2023	25/02/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	25/02/2023	25/02/2023
					CT981	Single use	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	25/02/2023	25/02/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	25/02/2023	25/02/2023
					CT983	Surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	25/02/2023	25/02/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	25/02/2023	25/02/2023
12493	2023		47265	Dual-chamber implantable pacemaker, rate-responsive	CT141	Electrodes/Leads and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	25/02/2023	25/02/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	25/02/2023	25/02/2023
					CT145	Cardiovascular devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	25/02/2023	25/02/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	25/02/2023	25/02/2023
					CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	25/02/2023	25/02/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	25/02/2023	25/02/2023
					CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	25/02/2023	25/02/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	25/02/2023	25/02/2023
					CT233	Surgical	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	25/02/2023	25/02/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	25/02/2023	25/02/2023
					CT2406	Implantable	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	25/02/2023	25/02/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	25/02/2023	25/02/2023
					CT2407	Active implantable	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	25/02/2023	25/02/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	25/02/2023	25/02/2023
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	25/02/2023	25/02/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	25/02/2023	25/02/2023
					CT2727	Active implant devices/ancillaries	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	25/02/2023	25/02/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	25/02/2023	25/02/2023
					CT2750	Implantable cardiac electrical stimulators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	25/02/2023	25/02/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	25/02/2023	25/02/2023
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	25/02/2023	25/02/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	25/02/2023	25/02/2023
					CT2985	Active	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	25/02/2023	25/02/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	25/02/2023	25/02/2023
					CT3001	Implantable metal	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	25/02/2023	25/02/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	25/02/2023	25/02/2023
					CT3002	Implantable synthetic polymer	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	25/02/2023	25/02/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	25/02/2023	25/02/2023
					CT3006	Implantable inorganic material	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	25/02/2023	25/02/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	25/02/2023	25/02/2023
					CT314	Electrically powered	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	25/02/2023	25/02/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	25/02/2023	25/02/2023
					CT321	Long-term surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	25/02/2023	25/02/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	25/02/2023	25/02/2023
					CT449	Generators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	25/02/2023	25/02/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	25/02/2023	25/02/2023
					CT981	Single use	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	25/02/2023	25/02/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	25/02/2023	25/02/2023
					CT983	Surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	25/02/2023	25/02/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	25/02/2023	25/02/2023
12494	2023		47265	Dual-chamber implantable pacemaker, rate-responsive	CT141	Electrodes/Leads and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	23/02/2023	23/02/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	23/02/2023	23/02/2023
					CT145	Cardiovascular devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	23/02/2023	23/02/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	23/02/2023	23/02/2023
					CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	23/02/2023	23/02/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	23/02/2023	23/02/2023
					CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	23/02/2023	23/02/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	23/02/2023	23/02/2023
					CT233	Surgical	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	23/02/2023	23/02/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	23/02/2023	23/02/2023
					CT2406	Implantable	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	23/02/2023	23/02/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	23/02/2023	23/02/2023
					CT2407	Active implantable	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	23/02/2023	23/02/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	23/02/2023	23/02/2023
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	23/02/2023	23/02/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	23/02/2023	23/02/2023

					CT2727	Active implant devices/ancillaries	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	23/02/2023	23/02/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	23/02/2023	23/02/2023
					CT2750	Implantable cardiac electrical stimulators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	23/02/2023	23/02/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	23/02/2023	23/02/2023
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	23/02/2023	23/02/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	23/02/2023	23/02/2023
					CT2985	Active	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	23/02/2023	23/02/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	23/02/2023	23/02/2023
					CT3001	Implantable metal	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	23/02/2023	23/02/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	23/02/2023	23/02/2023
					CT3002	Implantable synthetic polymer	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	23/02/2023	23/02/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	23/02/2023	23/02/2023
					CT3006	Implantable inorganic material	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	23/02/2023	23/02/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	23/02/2023	23/02/2023
					CT314	Electrically powered	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	23/02/2023	23/02/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	23/02/2023	23/02/2023
					CT321	Long-term surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	23/02/2023	23/02/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	23/02/2023	23/02/2023
					CT449	Generators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	23/02/2023	23/02/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	23/02/2023	23/02/2023
					CT981	Single use	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	23/02/2023	23/02/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	23/02/2023	23/02/2023
					CT983	Surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	23/02/2023	23/02/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	23/02/2023	23/02/2023
12495	2023		47265	Dual-chamber implantable pacemaker, rate-responsive	CT141	Electrodes/Leads and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	27/02/2023	27/02/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	27/02/2023	27/02/2023
					CT145	Cardiovascular devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	27/02/2023	27/02/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	27/02/2023	27/02/2023
					CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	27/02/2023	27/02/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	27/02/2023	27/02/2023
					CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	27/02/2023	27/02/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	27/02/2023	27/02/2023
					CT233	Surgical	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	27/02/2023	27/02/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	27/02/2023	27/02/2023
					CT2406	Implantable	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	27/02/2023	27/02/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	27/02/2023	27/02/2023
					CT2407	Active implantable	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	27/02/2023	27/02/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	27/02/2023	27/02/2023
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	27/02/2023	27/02/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	27/02/2023	27/02/2023
					CT2727	Active implant devices/ancillaries	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	27/02/2023	27/02/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	27/02/2023	27/02/2023
					CT2750	Implantable cardiac electrical stimulators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	27/02/2023	27/02/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	27/02/2023	27/02/2023
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	27/02/2023	27/02/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	27/02/2023	27/02/2023
					CT2985	Active	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	27/02/2023	27/02/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	27/02/2023	27/02/2023
					CT3001	Implantable metal	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	27/02/2023	27/02/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	27/02/2023	27/02/2023
					CT3002	Implantable synthetic polymer	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	27/02/2023	27/02/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	27/02/2023	27/02/2023
					CT3006	Implantable inorganic material	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	27/02/2023	27/02/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	27/02/2023	27/02/2023
					CT314	Electrically powered	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	27/02/2023	27/02/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	27/02/2023	27/02/2023
					CT321	Long-term surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	27/02/2023	27/02/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	27/02/2023	27/02/2023
					CT449	Generators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	27/02/2023	27/02/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	27/02/2023	27/02/2023
					CT981	Single use	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	27/02/2023	27/02/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	27/02/2023	27/02/2023
					CT983	Surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	27/02/2023	27/02/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	27/02/2023	27/02/2023
12496	2023	All other reportable incidents	17882	Physiologic-monitoring defibrillation system	CT1684	Cardiac pacemakers and associated devices	E0602	Cardiac Arrest	A071301	Failure to Deliver Shock/Stimulation		
							F05	Delay to Treatment/ Therapy	A071301	Failure to Deliver Shock/Stimulation		
12497	2023		47265	Dual-chamber implantable pacemaker, rate-responsive	CT141	Electrodes/Leads and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	08/02/2023	08/02/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	08/02/2023	08/02/2023
					CT145	Cardiovascular devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	08/02/2023	08/02/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	08/02/2023	08/02/2023
					CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	08/02/2023	08/02/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	08/02/2023	08/02/2023
					CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	08/02/2023	08/02/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	08/02/2023	08/02/2023
					CT233	Surgical	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	08/02/2023	08/02/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	08/02/2023	08/02/2023
					CT2406	Implantable	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	08/02/2023	08/02/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	08/02/2023	08/02/2023
					CT2407	Active implantable	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	08/02/2023	08/02/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	08/02/2023	08/02/2023
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	08/02/2023	08/02/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	08/02/2023	08/02/2023
					CT2727	Active implant devices/ancillaries	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	08/02/2023	08/02/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	08/02/2023	08/02/2023
					CT2750	Implantable cardiac electrical stimulators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	08/02/2023	08/02/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	08/02/2023	08/02/2023
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	08/02/2023	08/02/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	08/02/2023	08/02/2023

					CT2985	Active	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	08/02/2023	08/02/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	08/02/2023	08/02/2023
					CT3001	Implantable metal	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	08/02/2023	08/02/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	08/02/2023	08/02/2023
					CT3002	Implantable synthetic polymer	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	08/02/2023	08/02/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	08/02/2023	08/02/2023
					CT3006	Implantable inorganic material	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	08/02/2023	08/02/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	08/02/2023	08/02/2023
					CT314	Electrically powered	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	08/02/2023	08/02/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	08/02/2023	08/02/2023
					CT321	Long-term surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	08/02/2023	08/02/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	08/02/2023	08/02/2023
					CT449	Generators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	08/02/2023	08/02/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	08/02/2023	08/02/2023
					CT981	Single use	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	08/02/2023	08/02/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	08/02/2023	08/02/2023
					CT983	Surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	08/02/2023	08/02/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	08/02/2023	08/02/2023
12498	2023	All other reportable incidents	47265	Dual-chamber implantable pacemaker, rate-responsive	CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A070102	High Capture Threshold	22/12/2022	22/12/2022
									A070504	Premature Discharge of Battery	22/12/2022	22/12/2022
									A072201	High Impedance	22/12/2022	22/12/2022
							F08	Hospitalization or Prolonged Hospitalization	A070102	High Capture Threshold	22/12/2022	22/12/2022
									A070504	Premature Discharge of Battery	22/12/2022	22/12/2022
									A072201	High Impedance	22/12/2022	22/12/2022
							F1905	Device Revision or Replacement	A070102	High Capture Threshold	22/12/2022	22/12/2022
									A070504	Premature Discharge of Battery	22/12/2022	22/12/2022
									A072201	High Impedance	22/12/2022	22/12/2022
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A070102	High Capture Threshold	22/12/2022	22/12/2022
									A070504	Premature Discharge of Battery	22/12/2022	22/12/2022
									A072201	High Impedance	22/12/2022	22/12/2022
							F08	Hospitalization or Prolonged Hospitalization	A070102	High Capture Threshold	22/12/2022	22/12/2022
									A070504	Premature Discharge of Battery	22/12/2022	22/12/2022
									A072201	High Impedance	22/12/2022	22/12/2022
							F1905	Device Revision or Replacement	A070102	High Capture Threshold	22/12/2022	22/12/2022
									A070504	Premature Discharge of Battery	22/12/2022	22/12/2022
									A072201	High Impedance	22/12/2022	22/12/2022
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A070102	High Capture Threshold	22/12/2022	22/12/2022
									A070504	Premature Discharge of Battery	22/12/2022	22/12/2022
									A072201	High Impedance	22/12/2022	22/12/2022
							F08	Hospitalization or Prolonged Hospitalization	A070102	High Capture Threshold	22/12/2022	22/12/2022
									A070504	Premature Discharge of Battery	22/12/2022	22/12/2022
									A072201	High Impedance	22/12/2022	22/12/2022
							F1905	Device Revision or Replacement	A070102	High Capture Threshold	22/12/2022	22/12/2022
									A070504	Premature Discharge of Battery	22/12/2022	22/12/2022
									A072201	High Impedance	22/12/2022	22/12/2022
12499	2023	All other reportable incidents	35852	Single-chamber implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E2104	Electric Shock	A070909	Oversensing		
									A071302	Inappropriate/Inadequate Shock/Stimulation		
									A090801	Signal Artifact/Noise		
							F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A070909	Oversensing		
									A071302	Inappropriate/Inadequate Shock/Stimulation		
									A090801	Signal Artifact/Noise		
12500	2023	All other reportable incidents	35852	Single-chamber implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E2104	Electric Shock	A070909	Oversensing		
									A071302	Inappropriate/Inadequate Shock/Stimulation		
									A090801	Signal Artifact/Noise		
							F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A070909	Oversensing		
									A071302	Inappropriate/Inadequate Shock/Stimulation		
									A090801	Signal Artifact/Noise		
12501	2023	All other reportable incidents	37265	Dual-chamber implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
12502	2023	All other reportable incidents	35223	Endocardial/interventricular septal pacing lead	CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A0401	Break		
									A070909	Oversensing		
							F1901	Additional Surgery	A0401	Break		
									A070909	Oversensing		
12503	2023		47265	Dual-chamber implantable pacemaker, rate-responsive	CT141	Electrodes/Leads and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	17/02/2023	17/02/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	17/02/2023	17/02/2023
					CT145	Cardiovascular devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	17/02/2023	17/02/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	17/02/2023	17/02/2023
					CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	17/02/2023	17/02/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	17/02/2023	17/02/2023
					CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	17/02/2023	17/02/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	17/02/2023	17/02/2023
					CT233	Surgical	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	17/02/2023	17/02/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	17/02/2023	17/02/2023
					CT2406	Implantable	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	17/02/2023	17/02/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	17/02/2023	17/02/2023
					CT2407	Active implantable	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	17/02/2023	17/02/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	17/02/2023	17/02/2023
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	17/02/2023	17/02/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	17/02/2023	17/02/2023
					CT2727	Active implant devices/ancillaries	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	17/02/2023	17/02/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	17/02/2023	17/02/2023
					CT2750	Implantable cardiac electrical stimulators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	17/02/2023	17/02/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	17/02/2023	17/02/2023
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	17/02/2023	17/02/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	17/02/2023	17/02/2023
					CT2985	Active	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	17/02/2023	17/02/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	17/02/2023	17/02/2023
					CT3001	Implantable metal	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	17/02/2023	17/02/2023

							F1901	Additional Surgery	A24	No Apparent Device Problem	17/02/2023	17/02/2023
					CT3002	Implantable synthetic polymer	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	17/02/2023	17/02/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	17/02/2023	17/02/2023
					CT3006	Implantable inorganic material	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	17/02/2023	17/02/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	17/02/2023	17/02/2023
					CT314	Electrically powered	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	17/02/2023	17/02/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	17/02/2023	17/02/2023
					CT321	Long-term surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	17/02/2023	17/02/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	17/02/2023	17/02/2023
					CT449	Generators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	17/02/2023	17/02/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	17/02/2023	17/02/2023
					CT981	Single use	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	17/02/2023	17/02/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	17/02/2023	17/02/2023
					CT983	Surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	17/02/2023	17/02/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	17/02/2023	17/02/2023
12504	2023	All other reportable incidents	17882	Physiologic-monitoring defibrillation system	CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A12	Connection Problem		
							F26	No Health Consequences or Impact	A12	Connection Problem		
12505	2023	All other reportable incidents	35223	Endocardial/interventricular septal pacing lead	CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A072201	High Impedance	20/02/2023	20/02/2023
							F1901	Additional Surgery	A072201	High Impedance	20/02/2023	20/02/2023
12506	2023	All other reportable incidents	35223	Endocardial/interventricular septal pacing lead	CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A0701	Capturing Problem		
									A070102	High Capture Threshold		
									A0722	Impedance Problem		
									A072201	High Impedance		
									A2301	Device Handling Problem		
							F26	No Health Consequences or Impact	A0701	Capturing Problem		
									A070102	High Capture Threshold		
									A0722	Impedance Problem		
									A072201	High Impedance		
									A2301	Device Handling Problem		
12507	2023	All other reportable incidents	47265	Dual-chamber implantable pacemaker, rate-responsive	CT141	Electrodes/Leads and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A072202	Low Impedance		
									A090801	Signal Artifact/Noise		
							F1905	Device Revision or Replacement	A072202	Low Impedance		
									A090801	Signal Artifact/Noise		
							F1908	Prolonged Surgery	A072202	Low Impedance		
									A090801	Signal Artifact/Noise		
					CT145	Cardiovascular devices	E2403	No Clinical Signs, Symptoms or Conditions	A072202	Low Impedance		
									A090801	Signal Artifact/Noise		
							F1905	Device Revision or Replacement	A072202	Low Impedance		
									A090801	Signal Artifact/Noise		
							F1908	Prolonged Surgery	A072202	Low Impedance		
									A090801	Signal Artifact/Noise		
					CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A072202	Low Impedance		
									A090801	Signal Artifact/Noise		
							F1905	Device Revision or Replacement	A072202	Low Impedance		
									A090801	Signal Artifact/Noise		
							F1908	Prolonged Surgery	A072202	Low Impedance		
									A090801	Signal Artifact/Noise		
					CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A072202	Low Impedance		
									A090801	Signal Artifact/Noise		
							F1905	Device Revision or Replacement	A072202	Low Impedance		
									A090801	Signal Artifact/Noise		
							F1908	Prolonged Surgery	A072202	Low Impedance		
									A090801	Signal Artifact/Noise		
					CT233	Surgical	E2403	No Clinical Signs, Symptoms or Conditions	A072202	Low Impedance		
									A090801	Signal Artifact/Noise		
							F1905	Device Revision or Replacement	A072202	Low Impedance		
									A090801	Signal Artifact/Noise		
							F1908	Prolonged Surgery	A072202	Low Impedance		
									A090801	Signal Artifact/Noise		
					CT2406	Implantable	E2403	No Clinical Signs, Symptoms or Conditions	A072202	Low Impedance		
									A090801	Signal Artifact/Noise		
							F1905	Device Revision or Replacement	A072202	Low Impedance		
									A090801	Signal Artifact/Noise		
							F1908	Prolonged Surgery	A072202	Low Impedance		
									A090801	Signal Artifact/Noise		
					CT2407	Active implantable	E2403	No Clinical Signs, Symptoms or Conditions	A072202	Low Impedance		
									A090801	Signal Artifact/Noise		
							F1905	Device Revision or Replacement	A072202	Low Impedance		
									A090801	Signal Artifact/Noise		
							F1908	Prolonged Surgery	A072202	Low Impedance		
									A090801	Signal Artifact/Noise		
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A072202	Low Impedance		
									A090801	Signal Artifact/Noise		
							F1905	Device Revision or Replacement	A072202	Low Impedance		
									A090801	Signal Artifact/Noise		
							F1908	Prolonged Surgery	A072202	Low Impedance		
									A090801	Signal Artifact/Noise		
					CT2727	Active implant devices/ancillaries	E2403	No Clinical Signs, Symptoms or Conditions	A072202	Low Impedance		
									A090801	Signal Artifact/Noise		
							F1905	Device Revision or Replacement	A072202	Low Impedance		
									A090801	Signal Artifact/Noise		
							F1908	Prolonged Surgery	A072202	Low Impedance		
									A090801	Signal Artifact/Noise		
					CT2750	Implantable cardiac electrical stimulators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A072202	Low Impedance		
									A090801	Signal Artifact/Noise		
							F1905	Device Revision or Replacement	A072202	Low Impedance		
									A090801	Signal Artifact/Noise		
							F1908	Prolonged Surgery	A072202	Low Impedance		

									A090801	Signal Artifact/Noise			
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A072202	Low Impedance			
									A090801	Signal Artifact/Noise			
							F1905	Device Revision or Replacement	A072202	Low Impedance			
									A090801	Signal Artifact/Noise			
							F1908	Prolonged Surgery	A072202	Low Impedance			
									A090801	Signal Artifact/Noise			
					CT2985	Active	E2403	No Clinical Signs, Symptoms or Conditions	A072202	Low Impedance			
									A090801	Signal Artifact/Noise			
							F1905	Device Revision or Replacement	A072202	Low Impedance			
									A090801	Signal Artifact/Noise			
							F1908	Prolonged Surgery	A072202	Low Impedance			
									A090801	Signal Artifact/Noise			
					CT3001	Implantable metal	E2403	No Clinical Signs, Symptoms or Conditions	A072202	Low Impedance			
									A090801	Signal Artifact/Noise			
							F1905	Device Revision or Replacement	A072202	Low Impedance			
									A090801	Signal Artifact/Noise			
							F1908	Prolonged Surgery	A072202	Low Impedance			
									A090801	Signal Artifact/Noise			
					CT3002	Implantable synthetic polymer	E2403	No Clinical Signs, Symptoms or Conditions	A072202	Low Impedance			
									A090801	Signal Artifact/Noise			
							F1905	Device Revision or Replacement	A072202	Low Impedance			
									A090801	Signal Artifact/Noise			
							F1908	Prolonged Surgery	A072202	Low Impedance			
									A090801	Signal Artifact/Noise			
					CT3006	Implantable inorganic material	E2403	No Clinical Signs, Symptoms or Conditions	A072202	Low Impedance			
									A090801	Signal Artifact/Noise			
							F1905	Device Revision or Replacement	A072202	Low Impedance			
									A090801	Signal Artifact/Noise			
							F1908	Prolonged Surgery	A072202	Low Impedance			
									A090801	Signal Artifact/Noise			
					CT314	Electrically powered	E2403	No Clinical Signs, Symptoms or Conditions	A072202	Low Impedance			
									A090801	Signal Artifact/Noise			
							F1905	Device Revision or Replacement	A072202	Low Impedance			
									A090801	Signal Artifact/Noise			
							F1908	Prolonged Surgery	A072202	Low Impedance			
									A090801	Signal Artifact/Noise			
					CT321	Long-term surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A072202	Low Impedance			
									A090801	Signal Artifact/Noise			
							F1905	Device Revision or Replacement	A072202	Low Impedance			
									A090801	Signal Artifact/Noise			
							F1908	Prolonged Surgery	A072202	Low Impedance			
									A090801	Signal Artifact/Noise			
					CT449	Generators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A072202	Low Impedance			
									A090801	Signal Artifact/Noise			
							F1905	Device Revision or Replacement	A072202	Low Impedance			
									A090801	Signal Artifact/Noise			
							F1908	Prolonged Surgery	A072202	Low Impedance			
									A090801	Signal Artifact/Noise			
					CT981	Single use	E2403	No Clinical Signs, Symptoms or Conditions	A072202	Low Impedance			
									A090801	Signal Artifact/Noise			
							F1905	Device Revision or Replacement	A072202	Low Impedance			
									A090801	Signal Artifact/Noise			
							F1908	Prolonged Surgery	A072202	Low Impedance			
									A090801	Signal Artifact/Noise			
					CT983	Surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A072202	Low Impedance			
									A090801	Signal Artifact/Noise			
							F1905	Device Revision or Replacement	A072202	Low Impedance			
									A090801	Signal Artifact/Noise			
							F1908	Prolonged Surgery	A072202	Low Impedance			
									A090801	Signal Artifact/Noise			
12508	2023	All other reportable incidents	35852	Single-chamber implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem			
									A070504	Premature Discharge of Battery			
							F26	No Health Consequences or Impact	A0705	Battery Problem			
									A070504	Premature Discharge of Battery			
12509	2023		47265	Dual-chamber implantable pacemaker, rate-responsive	CT141	Electrodes/Leads and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	04/02/2023	04/02/2023	
							F1901	Additional Surgery	A24	No Apparent Device Problem	04/02/2023	04/02/2023	
					CT145	Cardiovascular devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	04/02/2023	04/02/2023	
							F1901	Additional Surgery	A24	No Apparent Device Problem	04/02/2023	04/02/2023	
					CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	04/02/2023	04/02/2023	
							F1901	Additional Surgery	A24	No Apparent Device Problem	04/02/2023	04/02/2023	
					CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	04/02/2023	04/02/2023	
							F1901	Additional Surgery	A24	No Apparent Device Problem	04/02/2023	04/02/2023	
					CT233	Surgical	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	04/02/2023	04/02/2023	
							F1901	Additional Surgery	A24	No Apparent Device Problem	04/02/2023	04/02/2023	
					CT2406	Implantable	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	04/02/2023	04/02/2023	
							F1901	Additional Surgery	A24	No Apparent Device Problem	04/02/2023	04/02/2023	
					CT2407	Active implantable	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	04/02/2023	04/02/2023	
							F1901	Additional Surgery	A24	No Apparent Device Problem	04/02/2023	04/02/2023	
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	04/02/2023	04/02/2023	
							F1901	Additional Surgery	A24	No Apparent Device Problem	04/02/2023	04/02/2023	
					CT2727	Active implant devices/ancillaries	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	04/02/2023	04/02/2023	
							F1901	Additional Surgery	A24	No Apparent Device Problem	04/02/2023	04/02/2023	
					CT2750	Implantable cardiac electrical stimulators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	04/02/2023	04/02/2023	
							F1901	Additional Surgery	A24	No Apparent Device Problem	04/02/2023	04/02/2023	
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	04/02/2023	04/02/2023	
							F1901	Additional Surgery	A24	No Apparent Device Problem	04/02/2023	04/02/2023	
					CT2985	Active	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	04/02/2023	04/02/2023	

							F1901	Additional Surgery	A24	No Apparent Device Problem	04/02/2023	04/02/2023
					CT3001	Implantable metal	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	04/02/2023	04/02/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	04/02/2023	04/02/2023
					CT3002	Implantable synthetic polymer	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	04/02/2023	04/02/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	04/02/2023	04/02/2023
					CT3006	Implantable inorganic material	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	04/02/2023	04/02/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	04/02/2023	04/02/2023
					CT314	Electrically powered	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	04/02/2023	04/02/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	04/02/2023	04/02/2023
					CT321	Long-term surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	04/02/2023	04/02/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	04/02/2023	04/02/2023
					CT449	Generators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	04/02/2023	04/02/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	04/02/2023	04/02/2023
					CT981	Single use	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	04/02/2023	04/02/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	04/02/2023	04/02/2023
					CT983	Surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	04/02/2023	04/02/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	04/02/2023	04/02/2023
12510	2023	All other reportable incidents	35852	Single-chamber implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A090801	Signal Artifact/Noise		
									A1304	Telemetry Discrepancy		
							F26	No Health Consequences or Impact	A090801	Signal Artifact/Noise		
									A1304	Telemetry Discrepancy		
12511	2023	All other reportable incidents	47270	Cardiac resynchronization therapy implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A072201	High Impedance		
							F26	No Health Consequences or Impact	A072201	High Impedance		
12512	2023	All other reportable incidents	17882	Physiologic-monitoring defibrillation system	CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A0719	Unexpected Shutdown		
							F27	Problem Identified During Non-Clinical Procedure	A0719	Unexpected Shutdown		
12513	2023	All other reportable incidents	35223	Endocardial/interventricular septal pacing lead	CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070909	Oversensing		
									A072201	High Impedance		
									A090801	Signal Artifact/Noise		
							F26	No Health Consequences or Impact	A070909	Oversensing		
									A072201	High Impedance		
									A090801	Signal Artifact/Noise		
12514	2023		12913	Pacemaker, implantable	CT145	Cardiovascular devices	F1903	Device Explantation	A070504	Premature Discharge of Battery		
					CT1684	Cardiac pacemakers and associated devices	F1903	Device Explantation	A070504	Premature Discharge of Battery		
					CT1685	Cardiac pacemakers	F1903	Device Explantation	A070504	Premature Discharge of Battery		
					CT2406	Implantable	F1903	Device Explantation	A070504	Premature Discharge of Battery		
					CT2407	Active implantable	F1903	Device Explantation	A070504	Premature Discharge of Battery		
					CT257	Implantable cardiac pacemakers	F1903	Device Explantation	A070504	Premature Discharge of Battery		
					CT2727	Active implant devices/ancillaries	F1903	Device Explantation	A070504	Premature Discharge of Battery		
					CT2750	Implantable cardiac electrical stimulators and associated devices	F1903	Device Explantation	A070504	Premature Discharge of Battery		
					CT2866	Implantable cardiac electrical stimulators	F1903	Device Explantation	A070504	Premature Discharge of Battery		
					CT2985	Active	F1903	Device Explantation	A070504	Premature Discharge of Battery		
					CT314	Electrically powered	F1903	Device Explantation	A070504	Premature Discharge of Battery		
12515	2023	All other reportable incidents	35852	Single-chamber implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A12	Connection Problem	21/04/2016	21/04/2016
							F1908	Prolonged Surgery	A12	Connection Problem	21/04/2016	21/04/2016
12516	2023	All other reportable incidents	35852	Single-chamber implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A12	Connection Problem	25/09/2015	25/09/2015
							F1908	Prolonged Surgery	A12	Connection Problem	25/09/2015	25/09/2015
12517	2023	All other reportable incidents	60789	Intracardiac pacemaker	CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A070102	High Capture Threshold	30/12/2022	30/12/2022
							F12	Serious Injury/ Illness/ Impairment	A070102	High Capture Threshold	30/12/2022	30/12/2022
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A070102	High Capture Threshold	30/12/2022	30/12/2022
							F12	Serious Injury/ Illness/ Impairment	A070102	High Capture Threshold	30/12/2022	30/12/2022
					CT2818	Catheterization/Cannulation/Fluid path devices	E2403	No Clinical Signs, Symptoms or Conditions	A070102	High Capture Threshold	30/12/2022	30/12/2022
							F12	Serious Injury/ Illness/ Impairment	A070102	High Capture Threshold	30/12/2022	30/12/2022
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A070102	High Capture Threshold	30/12/2022	30/12/2022
							F12	Serious Injury/ Illness/ Impairment	A070102	High Capture Threshold	30/12/2022	30/12/2022
12518	2023		47265	Dual-chamber implantable pacemaker, rate-responsive	CT141	Electrodes/Leads and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	14/02/2023	14/02/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	14/02/2023	14/02/2023
					CT145	Cardiovascular devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	14/02/2023	14/02/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	14/02/2023	14/02/2023
					CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	14/02/2023	14/02/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	14/02/2023	14/02/2023
					CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	14/02/2023	14/02/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	14/02/2023	14/02/2023
					CT233	Surgical	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	14/02/2023	14/02/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	14/02/2023	14/02/2023
					CT2406	Implantable	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	14/02/2023	14/02/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	14/02/2023	14/02/2023
					CT2407	Active implantable	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	14/02/2023	14/02/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	14/02/2023	14/02/2023
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	14/02/2023	14/02/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	14/02/2023	14/02/2023
					CT2727	Active implant devices/ancillaries	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	14/02/2023	14/02/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	14/02/2023	14/02/2023
					CT2750	Implantable cardiac electrical stimulators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	14/02/2023	14/02/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	14/02/2023	14/02/2023
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	14/02/2023	14/02/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	14/02/2023	14/02/2023
					CT2985	Active	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	14/02/2023	14/02/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	14/02/2023	14/02/2023
					CT3001	Implantable metal	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	14/02/2023	14/02/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	14/02/2023	14/02/2023
					CT3002	Implantable synthetic polymer	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	14/02/2023	14/02/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	14/02/2023	14/02/2023
					CT3006	Implantable inorganic material	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	14/02/2023	14/02/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	14/02/2023	14/02/2023
					CT314	Electrically powered	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	14/02/2023	14/02/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	14/02/2023	14/02/2023
					CT321	Long-term surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	14/02/2023	14/02/2023

							F1901	Additional Surgery	A24	No Apparent Device Problem	14/02/2023	14/02/2023
					CT449	Generators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	14/02/2023	14/02/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	14/02/2023	14/02/2023
					CT981	Single use	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	14/02/2023	14/02/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	14/02/2023	14/02/2023
					CT983	Surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	14/02/2023	14/02/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	14/02/2023	14/02/2023
12519	2023	All other reportable incidents	17882	Physiologic-monitoring defibrillation system	CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A0708	Power Problem		
							F27	Problem Identified During Non-Clinical Procedure	A0708	Power Problem		
12520	2023		47265	Dual-chamber implantable pacemaker, rate-responsive	CT141	Electrodes/Leads and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	10/02/2023	10/02/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	10/02/2023	10/02/2023
					CT145	Cardiovascular devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	10/02/2023	10/02/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	10/02/2023	10/02/2023
					CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	10/02/2023	10/02/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	10/02/2023	10/02/2023
					CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	10/02/2023	10/02/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	10/02/2023	10/02/2023
					CT233	Surgical	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	10/02/2023	10/02/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	10/02/2023	10/02/2023
					CT2406	Implantable	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	10/02/2023	10/02/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	10/02/2023	10/02/2023
					CT2407	Active implantable	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	10/02/2023	10/02/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	10/02/2023	10/02/2023
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	10/02/2023	10/02/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	10/02/2023	10/02/2023
					CT2727	Active implant devices/ancillaries	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	10/02/2023	10/02/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	10/02/2023	10/02/2023
					CT2750	Implantable cardiac electrical stimulators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	10/02/2023	10/02/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	10/02/2023	10/02/2023
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	10/02/2023	10/02/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	10/02/2023	10/02/2023
					CT2985	Active	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	10/02/2023	10/02/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	10/02/2023	10/02/2023
					CT3001	Implantable metal	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	10/02/2023	10/02/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	10/02/2023	10/02/2023
					CT3002	Implantable synthetic polymer	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	10/02/2023	10/02/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	10/02/2023	10/02/2023
					CT3006	Implantable inorganic material	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	10/02/2023	10/02/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	10/02/2023	10/02/2023
					CT314	Electrically powered	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	10/02/2023	10/02/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	10/02/2023	10/02/2023
					CT321	Long-term surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	10/02/2023	10/02/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	10/02/2023	10/02/2023
					CT449	Generators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	10/02/2023	10/02/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	10/02/2023	10/02/2023
					CT981	Single use	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	10/02/2023	10/02/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	10/02/2023	10/02/2023
					CT983	Surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	10/02/2023	10/02/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	10/02/2023	10/02/2023
12521	2023		47265	Dual-chamber implantable pacemaker, rate-responsive	CT141	Electrodes/Leads and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	28/02/2023	28/02/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	28/02/2023	28/02/2023
					CT145	Cardiovascular devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	28/02/2023	28/02/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	28/02/2023	28/02/2023
					CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	28/02/2023	28/02/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	28/02/2023	28/02/2023
					CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	28/02/2023	28/02/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	28/02/2023	28/02/2023
					CT233	Surgical	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	28/02/2023	28/02/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	28/02/2023	28/02/2023
					CT2406	Implantable	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	28/02/2023	28/02/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	28/02/2023	28/02/2023
					CT2407	Active implantable	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	28/02/2023	28/02/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	28/02/2023	28/02/2023
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	28/02/2023	28/02/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	28/02/2023	28/02/2023
					CT2727	Active implant devices/ancillaries	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	28/02/2023	28/02/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	28/02/2023	28/02/2023
					CT2750	Implantable cardiac electrical stimulators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	28/02/2023	28/02/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	28/02/2023	28/02/2023
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	28/02/2023	28/02/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	28/02/2023	28/02/2023
					CT2985	Active	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	28/02/2023	28/02/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	28/02/2023	28/02/2023
					CT3001	Implantable metal	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	28/02/2023	28/02/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	28/02/2023	28/02/2023
					CT3002	Implantable synthetic polymer	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	28/02/2023	28/02/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	28/02/2023	28/02/2023
					CT3006	Implantable inorganic material	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	28/02/2023	28/02/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	28/02/2023	28/02/2023
					CT314	Electrically powered	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	28/02/2023	28/02/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	28/02/2023	28/02/2023
					CT321	Long-term surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	28/02/2023	28/02/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	28/02/2023	28/02/2023
					CT449	Generators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	28/02/2023	28/02/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	28/02/2023	28/02/2023
					CT981	Single use	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	28/02/2023	28/02/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	28/02/2023	28/02/2023
					CT983	Surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	28/02/2023	28/02/2023

							F1901	Additional Surgery	A24	No Apparent Device Problem	28/02/2023	28/02/2023
12522	2023		47265	Dual-chamber implantable pacemaker, rate-responsive	CT141	Electrodes/Leads and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	28/02/2023	28/02/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	28/02/2023	28/02/2023
					CT145	Cardiovascular devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	28/02/2023	28/02/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	28/02/2023	28/02/2023
					CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	28/02/2023	28/02/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	28/02/2023	28/02/2023
					CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	28/02/2023	28/02/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	28/02/2023	28/02/2023
					CT233	Surgical	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	28/02/2023	28/02/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	28/02/2023	28/02/2023
					CT2406	Implantable	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	28/02/2023	28/02/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	28/02/2023	28/02/2023
					CT2407	Active implantable	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	28/02/2023	28/02/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	28/02/2023	28/02/2023
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	28/02/2023	28/02/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	28/02/2023	28/02/2023
					CT2727	Active implant devices/ancillaries	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	28/02/2023	28/02/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	28/02/2023	28/02/2023
					CT2750	Implantable cardiac electrical stimulators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	28/02/2023	28/02/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	28/02/2023	28/02/2023
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	28/02/2023	28/02/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	28/02/2023	28/02/2023
					CT2985	Active	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	28/02/2023	28/02/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	28/02/2023	28/02/2023
					CT3001	Implantable metal	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	28/02/2023	28/02/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	28/02/2023	28/02/2023
					CT3002	Implantable synthetic polymer	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	28/02/2023	28/02/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	28/02/2023	28/02/2023
					CT3006	Implantable inorganic material	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	28/02/2023	28/02/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	28/02/2023	28/02/2023
					CT314	Electrically powered	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	28/02/2023	28/02/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	28/02/2023	28/02/2023
					CT321	Long-term surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	28/02/2023	28/02/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	28/02/2023	28/02/2023
					CT449	Generators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	28/02/2023	28/02/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	28/02/2023	28/02/2023
					CT981	Single use	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	28/02/2023	28/02/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	28/02/2023	28/02/2023
					CT983	Surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	28/02/2023	28/02/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	28/02/2023	28/02/2023
12523	2023	All other reportable incidents	47267	Single-chamber implantable pacemaker, rate-responsive	CT141	Electrodes/Leads and associated devices	E060104	Bradycardia	A070103	Intermittent Capture	23/02/2023	23/02/2023
							F1903	Device Explantation	A070103	Intermittent Capture	23/02/2023	23/02/2023
					CT145	Cardiovascular devices	E060104	Bradycardia	A070103	Intermittent Capture	23/02/2023	23/02/2023
							F1903	Device Explantation	A070103	Intermittent Capture	23/02/2023	23/02/2023
					CT1684	Cardiac pacemakers and associated devices	E060104	Bradycardia	A070103	Intermittent Capture	23/02/2023	23/02/2023
							F1903	Device Explantation	A070103	Intermittent Capture	23/02/2023	23/02/2023
					CT1685	Cardiac pacemakers	E060104	Bradycardia	A070103	Intermittent Capture	23/02/2023	23/02/2023
							F1903	Device Explantation	A070103	Intermittent Capture	23/02/2023	23/02/2023
					CT233	Surgical	E060104	Bradycardia	A070103	Intermittent Capture	23/02/2023	23/02/2023
							F1903	Device Explantation	A070103	Intermittent Capture	23/02/2023	23/02/2023
					CT2406	Implantable	E060104	Bradycardia	A070103	Intermittent Capture	23/02/2023	23/02/2023
							F1903	Device Explantation	A070103	Intermittent Capture	23/02/2023	23/02/2023
					CT2407	Active implantable	E060104	Bradycardia	A070103	Intermittent Capture	23/02/2023	23/02/2023
							F1903	Device Explantation	A070103	Intermittent Capture	23/02/2023	23/02/2023
					CT257	Implantable cardiac pacemakers	E060104	Bradycardia	A070103	Intermittent Capture	23/02/2023	23/02/2023
							F1903	Device Explantation	A070103	Intermittent Capture	23/02/2023	23/02/2023
					CT2727	Active implant devices/ancillaries	E060104	Bradycardia	A070103	Intermittent Capture	23/02/2023	23/02/2023
							F1903	Device Explantation	A070103	Intermittent Capture	23/02/2023	23/02/2023
					CT2750	Implantable cardiac electrical stimulators and associated devices	E060104	Bradycardia	A070103	Intermittent Capture	23/02/2023	23/02/2023
							F1903	Device Explantation	A070103	Intermittent Capture	23/02/2023	23/02/2023
					CT2866	Implantable cardiac electrical stimulators	E060104	Bradycardia	A070103	Intermittent Capture	23/02/2023	23/02/2023
							F1903	Device Explantation	A070103	Intermittent Capture	23/02/2023	23/02/2023
					CT2985	Active	E060104	Bradycardia	A070103	Intermittent Capture	23/02/2023	23/02/2023
							F1903	Device Explantation	A070103	Intermittent Capture	23/02/2023	23/02/2023
					CT3001	Implantable metal	E060104	Bradycardia	A070103	Intermittent Capture	23/02/2023	23/02/2023
							F1903	Device Explantation	A070103	Intermittent Capture	23/02/2023	23/02/2023
					CT3002	Implantable synthetic polymer	E060104	Bradycardia	A070103	Intermittent Capture	23/02/2023	23/02/2023
							F1903	Device Explantation	A070103	Intermittent Capture	23/02/2023	23/02/2023
					CT3006	Implantable inorganic material	E060104	Bradycardia	A070103	Intermittent Capture	23/02/2023	23/02/2023
							F1903	Device Explantation	A070103	Intermittent Capture	23/02/2023	23/02/2023
					CT314	Electrically powered	E060104	Bradycardia	A070103	Intermittent Capture	23/02/2023	23/02/2023
							F1903	Device Explantation	A070103	Intermittent Capture	23/02/2023	23/02/2023
					CT321	Long-term surgical invasive	E060104	Bradycardia	A070103	Intermittent Capture	23/02/2023	23/02/2023
							F1903	Device Explantation	A070103	Intermittent Capture	23/02/2023	23/02/2023
					CT449	Generators and associated devices	E060104	Bradycardia	A070103	Intermittent Capture	23/02/2023	23/02/2023
							F1903	Device Explantation	A070103	Intermittent Capture	23/02/2023	23/02/2023
					CT981	Single use	E060104	Bradycardia	A070103	Intermittent Capture	23/02/2023	23/02/2023
							F1903	Device Explantation	A070103	Intermittent Capture	23/02/2023	23/02/2023
					CT983	Surgical invasive	E060104	Bradycardia	A070103	Intermittent Capture	23/02/2023	23/02/2023
							F1903	Device Explantation	A070103	Intermittent Capture	23/02/2023	23/02/2023
12524	2023	All other reportable incidents	35223	Endocardial/interventricular septal pacing lead	CT1684	Cardiac pacemakers and associated devices	E060104	Bradycardia	A070101	Failure to Capture	23/02/2023	23/02/2023
							F1903	Device Explantation	A070101	Failure to Capture	23/02/2023	23/02/2023
12525	2023	All other reportable incidents	35852	Single-chamber implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E060109	Tachycardia	A160602	Premature End-of-Life Indicator		
							F08	Hospitalization or Prolonged Hospitalization	A160602	Premature End-of-Life Indicator		
12526	2023	All other reportable incidents	35039	Epicardial pacing lead	CT1684	Cardiac pacemakers and associated devices	E060103	Atrial Flutter	A070909	Oversensing		
							F11	Non-Serious Injury/ Illness/ Impairment	A070909	Oversensing		
							F12	Serious Injury/ Illness/ Impairment	A070909	Oversensing		

							F23	Unexpected Medical Intervention	A070909	Oversensing			
12527	2023	All other reportable incidents	47265	Dual-chamber implantable pacemaker, rate-responsive	CT141	Electrodes/Leads and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	26/01/2023	26/01/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	26/01/2023	26/01/2023
					CT145	Cardiovascular devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	26/01/2023	26/01/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	26/01/2023	26/01/2023
					CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	26/01/2023	26/01/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	26/01/2023	26/01/2023
					CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	26/01/2023	26/01/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	26/01/2023	26/01/2023
					CT233	Surgical	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	26/01/2023	26/01/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	26/01/2023	26/01/2023
					CT2406	Implantable	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	26/01/2023	26/01/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	26/01/2023	26/01/2023
					CT2407	Active implantable	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	26/01/2023	26/01/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	26/01/2023	26/01/2023
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	26/01/2023	26/01/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	26/01/2023	26/01/2023
					CT2727	Active implant devices/ancillaries	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	26/01/2023	26/01/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	26/01/2023	26/01/2023
					CT2750	Implantable cardiac electrical stimulators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	26/01/2023	26/01/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	26/01/2023	26/01/2023
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	26/01/2023	26/01/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	26/01/2023	26/01/2023
					CT2985	Active	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	26/01/2023	26/01/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	26/01/2023	26/01/2023
					CT3001	Implantable metal	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	26/01/2023	26/01/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	26/01/2023	26/01/2023
					CT3002	Implantable synthetic polymer	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	26/01/2023	26/01/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	26/01/2023	26/01/2023
					CT3006	Implantable inorganic material	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	26/01/2023	26/01/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	26/01/2023	26/01/2023
					CT314	Electrically powered	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	26/01/2023	26/01/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	26/01/2023	26/01/2023
					CT321	Long-term surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	26/01/2023	26/01/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	26/01/2023	26/01/2023
					CT449	Generators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	26/01/2023	26/01/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	26/01/2023	26/01/2023
					CT981	Single use	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	26/01/2023	26/01/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	26/01/2023	26/01/2023
					CT983	Surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	26/01/2023	26/01/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	26/01/2023	26/01/2023
12528	2023		47265	Dual-chamber implantable pacemaker, rate-responsive	CT141	Electrodes/Leads and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	16/02/2023	16/02/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	16/02/2023	16/02/2023
					CT145	Cardiovascular devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	16/02/2023	16/02/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	16/02/2023	16/02/2023
					CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	16/02/2023	16/02/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	16/02/2023	16/02/2023
					CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	16/02/2023	16/02/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	16/02/2023	16/02/2023
					CT233	Surgical	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	16/02/2023	16/02/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	16/02/2023	16/02/2023
					CT2406	Implantable	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	16/02/2023	16/02/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	16/02/2023	16/02/2023
					CT2407	Active implantable	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	16/02/2023	16/02/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	16/02/2023	16/02/2023
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	16/02/2023	16/02/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	16/02/2023	16/02/2023
					CT2727	Active implant devices/ancillaries	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	16/02/2023	16/02/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	16/02/2023	16/02/2023
					CT2750	Implantable cardiac electrical stimulators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	16/02/2023	16/02/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	16/02/2023	16/02/2023
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	16/02/2023	16/02/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	16/02/2023	16/02/2023
					CT2985	Active	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	16/02/2023	16/02/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	16/02/2023	16/02/2023
					CT3001	Implantable metal	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	16/02/2023	16/02/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	16/02/2023	16/02/2023
					CT3002	Implantable synthetic polymer	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	16/02/2023	16/02/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	16/02/2023	16/02/2023
					CT3006	Implantable inorganic material	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	16/02/2023	16/02/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	16/02/2023	16/02/2023
					CT314	Electrically powered	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	16/02/2023	16/02/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	16/02/2023	16/02/2023
					CT321	Long-term surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	16/02/2023	16/02/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	16/02/2023	16/02/2023
					CT449	Generators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	16/02/2023	16/02/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	16/02/2023	16/02/2023
					CT981	Single use	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	16/02/2023	16/02/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	16/02/2023	16/02/2023
					CT983	Surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	16/02/2023	16/02/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	16/02/2023	16/02/2023
12529	2023		47265	Dual-chamber implantable pacemaker, rate-responsive	CT141	Electrodes/Leads and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	11/02/2023	11/02/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	11/02/2023	11/02/2023
					CT145	Cardiovascular devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	11/02/2023	11/02/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	11/02/2023	11/02/2023
					CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	11/02/2023	11/02/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	11/02/2023	11/02/2023
					CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	11/02/2023	11/02/2023

							F1901	Additional Surgery	A24	No Apparent Device Problem	11/02/2023	11/02/2023
					CT233	Surgical	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	11/02/2023	11/02/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	11/02/2023	11/02/2023
					CT2406	Implantable	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	11/02/2023	11/02/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	11/02/2023	11/02/2023
					CT2407	Active implantable	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	11/02/2023	11/02/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	11/02/2023	11/02/2023
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	11/02/2023	11/02/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	11/02/2023	11/02/2023
					CT2727	Active implant devices/ancillaries	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	11/02/2023	11/02/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	11/02/2023	11/02/2023
					CT2750	Implantable cardiac electrical stimulators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	11/02/2023	11/02/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	11/02/2023	11/02/2023
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	11/02/2023	11/02/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	11/02/2023	11/02/2023
					CT2985	Active	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	11/02/2023	11/02/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	11/02/2023	11/02/2023
					CT3001	Implantable metal	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	11/02/2023	11/02/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	11/02/2023	11/02/2023
					CT3002	Implantable synthetic polymer	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	11/02/2023	11/02/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	11/02/2023	11/02/2023
					CT3006	Implantable inorganic material	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	11/02/2023	11/02/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	11/02/2023	11/02/2023
					CT314	Electrically powered	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	11/02/2023	11/02/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	11/02/2023	11/02/2023
					CT321	Long-term surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	11/02/2023	11/02/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	11/02/2023	11/02/2023
					CT449	Generators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	11/02/2023	11/02/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	11/02/2023	11/02/2023
					CT981	Single use	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	11/02/2023	11/02/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	11/02/2023	11/02/2023
					CT983	Surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	11/02/2023	11/02/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	11/02/2023	11/02/2023
12530	2023		47265	Dual-chamber implantable pacemaker, rate-responsive	CT141	Electrodes/Leads and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	16/02/2023	16/02/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	16/02/2023	16/02/2023
					CT145	Cardiovascular devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	16/02/2023	16/02/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	16/02/2023	16/02/2023
					CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	16/02/2023	16/02/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	16/02/2023	16/02/2023
					CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	16/02/2023	16/02/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	16/02/2023	16/02/2023
					CT233	Surgical	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	16/02/2023	16/02/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	16/02/2023	16/02/2023
					CT2406	Implantable	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	16/02/2023	16/02/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	16/02/2023	16/02/2023
					CT2407	Active implantable	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	16/02/2023	16/02/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	16/02/2023	16/02/2023
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	16/02/2023	16/02/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	16/02/2023	16/02/2023
					CT2727	Active implant devices/ancillaries	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	16/02/2023	16/02/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	16/02/2023	16/02/2023
					CT2750	Implantable cardiac electrical stimulators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	16/02/2023	16/02/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	16/02/2023	16/02/2023
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	16/02/2023	16/02/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	16/02/2023	16/02/2023
					CT2985	Active	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	16/02/2023	16/02/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	16/02/2023	16/02/2023
					CT3001	Implantable metal	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	16/02/2023	16/02/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	16/02/2023	16/02/2023
					CT3002	Implantable synthetic polymer	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	16/02/2023	16/02/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	16/02/2023	16/02/2023
					CT3006	Implantable inorganic material	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	16/02/2023	16/02/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	16/02/2023	16/02/2023
					CT314	Electrically powered	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	16/02/2023	16/02/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	16/02/2023	16/02/2023
					CT321	Long-term surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	16/02/2023	16/02/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	16/02/2023	16/02/2023
					CT449	Generators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	16/02/2023	16/02/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	16/02/2023	16/02/2023
					CT981	Single use	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	16/02/2023	16/02/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	16/02/2023	16/02/2023
					CT983	Surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	16/02/2023	16/02/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	16/02/2023	16/02/2023
12531	2023	All other reportable incidents	47263	Cardiac resynchronization therapy implantable pacemaker	CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A0908	Incorrect, Inadequate or Imprecise Result or Readings		
							F26	No Health Consequences or Impact	A0908	Incorrect, Inadequate or Imprecise Result or Readings		
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A0908	Incorrect, Inadequate or Imprecise Result or Readings		
							F26	No Health Consequences or Impact	A0908	Incorrect, Inadequate or Imprecise Result or Readings		
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A0908	Incorrect, Inadequate or Imprecise Result or Readings		
							F26	No Health Consequences or Impact	A0908	Incorrect, Inadequate or Imprecise Result or Readings		
12532	2023		47265	Dual-chamber implantable pacemaker, rate-responsive	CT141	Electrodes/Leads and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	01/02/2023	01/02/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	01/02/2023	01/02/2023
					CT145	Cardiovascular devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	01/02/2023	01/02/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	01/02/2023	01/02/2023

					CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	01/02/2023	01/02/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	01/02/2023	01/02/2023
					CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	01/02/2023	01/02/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	01/02/2023	01/02/2023
					CT233	Surgical	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	01/02/2023	01/02/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	01/02/2023	01/02/2023
					CT2406	Implantable	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	01/02/2023	01/02/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	01/02/2023	01/02/2023
					CT2407	Active implantable	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	01/02/2023	01/02/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	01/02/2023	01/02/2023
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	01/02/2023	01/02/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	01/02/2023	01/02/2023
					CT2727	Active implant devices/ancillaries	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	01/02/2023	01/02/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	01/02/2023	01/02/2023
					CT2750	Implantable cardiac electrical stimulators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	01/02/2023	01/02/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	01/02/2023	01/02/2023
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	01/02/2023	01/02/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	01/02/2023	01/02/2023
					CT2985	Active	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	01/02/2023	01/02/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	01/02/2023	01/02/2023
					CT3001	Implantable metal	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	01/02/2023	01/02/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	01/02/2023	01/02/2023
					CT3002	Implantable synthetic polymer	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	01/02/2023	01/02/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	01/02/2023	01/02/2023
					CT3006	Implantable inorganic material	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	01/02/2023	01/02/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	01/02/2023	01/02/2023
					CT314	Electrically powered	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	01/02/2023	01/02/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	01/02/2023	01/02/2023
					CT321	Long-term surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	01/02/2023	01/02/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	01/02/2023	01/02/2023
					CT449	Generators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	01/02/2023	01/02/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	01/02/2023	01/02/2023
					CT981	Single use	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	01/02/2023	01/02/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	01/02/2023	01/02/2023
					CT983	Surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	01/02/2023	01/02/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	01/02/2023	01/02/2023
12533	2023		47265	Dual-chamber implantable pacemaker, rate-responsive	CT141	Electrodes/Leads and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	20/02/2023	20/02/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	20/02/2023	20/02/2023
					CT145	Cardiovascular devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	20/02/2023	20/02/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	20/02/2023	20/02/2023
					CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	20/02/2023	20/02/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	20/02/2023	20/02/2023
					CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	20/02/2023	20/02/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	20/02/2023	20/02/2023
					CT233	Surgical	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	20/02/2023	20/02/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	20/02/2023	20/02/2023
					CT2406	Implantable	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	20/02/2023	20/02/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	20/02/2023	20/02/2023
					CT2407	Active implantable	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	20/02/2023	20/02/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	20/02/2023	20/02/2023
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	20/02/2023	20/02/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	20/02/2023	20/02/2023
					CT2727	Active implant devices/ancillaries	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	20/02/2023	20/02/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	20/02/2023	20/02/2023
					CT2750	Implantable cardiac electrical stimulators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	20/02/2023	20/02/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	20/02/2023	20/02/2023
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	20/02/2023	20/02/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	20/02/2023	20/02/2023
					CT2985	Active	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	20/02/2023	20/02/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	20/02/2023	20/02/2023
					CT3001	Implantable metal	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	20/02/2023	20/02/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	20/02/2023	20/02/2023
					CT3002	Implantable synthetic polymer	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	20/02/2023	20/02/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	20/02/2023	20/02/2023
					CT3006	Implantable inorganic material	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	20/02/2023	20/02/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	20/02/2023	20/02/2023
					CT314	Electrically powered	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	20/02/2023	20/02/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	20/02/2023	20/02/2023
					CT321	Long-term surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	20/02/2023	20/02/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	20/02/2023	20/02/2023
					CT449	Generators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	20/02/2023	20/02/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	20/02/2023	20/02/2023
					CT981	Single use	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	20/02/2023	20/02/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	20/02/2023	20/02/2023
					CT983	Surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	20/02/2023	20/02/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	20/02/2023	20/02/2023
12534	2023	All other reportable incidents	37265	Dual-chamber implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E2104	Electric Shock	A070909	Oversensing		
									A071302	Inappropriate/Inadequate Shock/Stimulation		
							F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A070909	Oversensing		
									A071302	Inappropriate/Inadequate Shock/Stimulation		
12535	2023	All other reportable incidents	47265	Dual-chamber implantable pacemaker, rate-responsive	CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		

					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
12536	2023		47265	Dual-chamber implantable pacemaker, rate-responsive	CT141	Electrodes/Leads and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	04/03/2023	04/03/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	04/03/2023	04/03/2023
					CT145	Cardiovascular devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	04/03/2023	04/03/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	04/03/2023	04/03/2023
					CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	04/03/2023	04/03/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	04/03/2023	04/03/2023
					CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	04/03/2023	04/03/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	04/03/2023	04/03/2023
					CT233	Surgical	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	04/03/2023	04/03/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	04/03/2023	04/03/2023
					CT2406	Implantable	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	04/03/2023	04/03/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	04/03/2023	04/03/2023
					CT2407	Active implantable	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	04/03/2023	04/03/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	04/03/2023	04/03/2023
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	04/03/2023	04/03/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	04/03/2023	04/03/2023
					CT2727	Active implant devices/ancillaries	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	04/03/2023	04/03/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	04/03/2023	04/03/2023
					CT2750	Implantable cardiac electrical stimulators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	04/03/2023	04/03/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	04/03/2023	04/03/2023
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	04/03/2023	04/03/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	04/03/2023	04/03/2023
					CT2985	Active	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	04/03/2023	04/03/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	04/03/2023	04/03/2023
					CT3001	Implantable metal	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	04/03/2023	04/03/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	04/03/2023	04/03/2023
					CT3002	Implantable synthetic polymer	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	04/03/2023	04/03/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	04/03/2023	04/03/2023
					CT3006	Implantable inorganic material	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	04/03/2023	04/03/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	04/03/2023	04/03/2023
					CT314	Electrically powered	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	04/03/2023	04/03/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	04/03/2023	04/03/2023
					CT321	Long-term surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	04/03/2023	04/03/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	04/03/2023	04/03/2023
					CT449	Generators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	04/03/2023	04/03/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	04/03/2023	04/03/2023
					CT981	Single use	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	04/03/2023	04/03/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	04/03/2023	04/03/2023
					CT983	Surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	04/03/2023	04/03/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	04/03/2023	04/03/2023
12537	2023	All other reportable incidents			CT257	Implantable cardiac pacemakers	E2104	Electric Shock	A070909	Oversensing	28/04/2023	28/04/2023
									A071302	Inappropriate/Inadequate Shock/Stimulation	28/04/2023	28/04/2023
							F08	Hospitalization or Prolonged Hospitalization	A070909	Oversensing	28/04/2023	28/04/2023
									A071302	Inappropriate/Inadequate Shock/Stimulation	28/04/2023	28/04/2023
							F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A070909	Oversensing	28/04/2023	28/04/2023
									A071302	Inappropriate/Inadequate Shock/Stimulation	28/04/2023	28/04/2023
							F1905	Device Revision or Replacement	A070909	Oversensing	28/04/2023	28/04/2023
									A071302	Inappropriate/Inadequate Shock/Stimulation	28/04/2023	28/04/2023
12538	2023	All other reportable incidents	37265	Dual-chamber implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A072201	High Impedance		
							F26	No Health Consequences or Impact	A072201	High Impedance		
12539	2023	All other reportable incidents	47265	Dual-chamber implantable pacemaker, rate-responsive	CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
12540	2023	All other reportable incidents	35852	Single-chamber implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E060104	Bradycardia	A070909	Oversensing		
									A0712	Pacing Problem		
									A090801	Signal Artifact/Noise		
							F1001	Absence of Treatment	A070909	Oversensing		
									A0712	Pacing Problem		
									A090801	Signal Artifact/Noise		
12541	2023		47265	Dual-chamber implantable pacemaker, rate-responsive	CT141	Electrodes/Leads and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	04/03/2023	04/03/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	04/03/2023	04/03/2023
					CT145	Cardiovascular devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	04/03/2023	04/03/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	04/03/2023	04/03/2023
					CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	04/03/2023	04/03/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	04/03/2023	04/03/2023
					CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	04/03/2023	04/03/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	04/03/2023	04/03/2023
					CT233	Surgical	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	04/03/2023	04/03/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	04/03/2023	04/03/2023
					CT2406	Implantable	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	04/03/2023	04/03/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	04/03/2023	04/03/2023
					CT2407	Active implantable	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	04/03/2023	04/03/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	04/03/2023	04/03/2023
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	04/03/2023	04/03/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	04/03/2023	04/03/2023

					CT2727	Active implant devices/ancillaries	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	04/03/2023	04/03/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	04/03/2023	04/03/2023
					CT2750	Implantable cardiac electrical stimulators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	04/03/2023	04/03/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	04/03/2023	04/03/2023
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	04/03/2023	04/03/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	04/03/2023	04/03/2023
					CT2985	Active	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	04/03/2023	04/03/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	04/03/2023	04/03/2023
					CT3001	Implantable metal	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	04/03/2023	04/03/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	04/03/2023	04/03/2023
					CT3002	Implantable synthetic polymer	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	04/03/2023	04/03/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	04/03/2023	04/03/2023
					CT3006	Implantable inorganic material	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	04/03/2023	04/03/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	04/03/2023	04/03/2023
					CT314	Electrically powered	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	04/03/2023	04/03/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	04/03/2023	04/03/2023
					CT321	Long-term surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	04/03/2023	04/03/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	04/03/2023	04/03/2023
					CT449	Generators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	04/03/2023	04/03/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	04/03/2023	04/03/2023
					CT981	Single use	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	04/03/2023	04/03/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	04/03/2023	04/03/2023
					CT983	Surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	04/03/2023	04/03/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	04/03/2023	04/03/2023
12542	2023		47265	Dual-chamber implantable pacemaker, rate-responsive	CT141	Electrodes/Leads and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	29/01/2023	29/01/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	29/01/2023	29/01/2023
					CT145	Cardiovascular devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	29/01/2023	29/01/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	29/01/2023	29/01/2023
					CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	29/01/2023	29/01/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	29/01/2023	29/01/2023
					CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	29/01/2023	29/01/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	29/01/2023	29/01/2023
					CT233	Surgical	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	29/01/2023	29/01/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	29/01/2023	29/01/2023
					CT2406	Implantable	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	29/01/2023	29/01/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	29/01/2023	29/01/2023
					CT2407	Active implantable	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	29/01/2023	29/01/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	29/01/2023	29/01/2023
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	29/01/2023	29/01/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	29/01/2023	29/01/2023
					CT2727	Active implant devices/ancillaries	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	29/01/2023	29/01/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	29/01/2023	29/01/2023
					CT2750	Implantable cardiac electrical stimulators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	29/01/2023	29/01/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	29/01/2023	29/01/2023
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	29/01/2023	29/01/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	29/01/2023	29/01/2023
					CT2985	Active	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	29/01/2023	29/01/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	29/01/2023	29/01/2023
					CT3001	Implantable metal	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	29/01/2023	29/01/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	29/01/2023	29/01/2023
					CT3002	Implantable synthetic polymer	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	29/01/2023	29/01/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	29/01/2023	29/01/2023
					CT3006	Implantable inorganic material	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	29/01/2023	29/01/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	29/01/2023	29/01/2023
					CT314	Electrically powered	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	29/01/2023	29/01/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	29/01/2023	29/01/2023
					CT321	Long-term surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	29/01/2023	29/01/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	29/01/2023	29/01/2023
					CT449	Generators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	29/01/2023	29/01/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	29/01/2023	29/01/2023
					CT981	Single use	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	29/01/2023	29/01/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	29/01/2023	29/01/2023
					CT983	Surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	29/01/2023	29/01/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	29/01/2023	29/01/2023
12543	2023		47265	Dual-chamber implantable pacemaker, rate-responsive	CT141	Electrodes/Leads and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	13/02/2023	13/02/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	13/02/2023	13/02/2023
					CT145	Cardiovascular devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	13/02/2023	13/02/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	13/02/2023	13/02/2023
					CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	13/02/2023	13/02/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	13/02/2023	13/02/2023
					CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	13/02/2023	13/02/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	13/02/2023	13/02/2023
					CT233	Surgical	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	13/02/2023	13/02/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	13/02/2023	13/02/2023
					CT2406	Implantable	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	13/02/2023	13/02/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	13/02/2023	13/02/2023
					CT2407	Active implantable	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	13/02/2023	13/02/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	13/02/2023	13/02/2023
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	13/02/2023	13/02/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	13/02/2023	13/02/2023
					CT2727	Active implant devices/ancillaries	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	13/02/2023	13/02/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	13/02/2023	13/02/2023
					CT2750	Implantable cardiac electrical stimulators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	13/02/2023	13/02/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	13/02/2023	13/02/2023
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	13/02/2023	13/02/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	13/02/2023	13/02/2023
					CT2985	Active	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	13/02/2023	13/02/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	13/02/2023	13/02/2023

					CT3001	Implantable metal	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	13/02/2023	13/02/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	13/02/2023	13/02/2023
					CT3002	Implantable synthetic polymer	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	13/02/2023	13/02/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	13/02/2023	13/02/2023
					CT3006	Implantable inorganic material	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	13/02/2023	13/02/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	13/02/2023	13/02/2023
					CT314	Electrically powered	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	13/02/2023	13/02/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	13/02/2023	13/02/2023
					CT321	Long-term surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	13/02/2023	13/02/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	13/02/2023	13/02/2023
					CT449	Generators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	13/02/2023	13/02/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	13/02/2023	13/02/2023
					CT981	Single use	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	13/02/2023	13/02/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	13/02/2023	13/02/2023
					CT983	Surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	13/02/2023	13/02/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	13/02/2023	13/02/2023
12544	2023		47265	Dual-chamber implantable pacemaker, rate-responsive	CT141	Electrodes/Leads and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	02/02/2023	02/02/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	02/02/2023	02/02/2023
					CT145	Cardiovascular devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	02/02/2023	02/02/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	02/02/2023	02/02/2023
					CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	02/02/2023	02/02/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	02/02/2023	02/02/2023
					CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	02/02/2023	02/02/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	02/02/2023	02/02/2023
					CT233	Surgical	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	02/02/2023	02/02/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	02/02/2023	02/02/2023
					CT2406	Implantable	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	02/02/2023	02/02/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	02/02/2023	02/02/2023
					CT2407	Active implantable	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	02/02/2023	02/02/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	02/02/2023	02/02/2023
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	02/02/2023	02/02/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	02/02/2023	02/02/2023
					CT2727	Active implant devices/ancillaries	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	02/02/2023	02/02/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	02/02/2023	02/02/2023
					CT2750	Implantable cardiac electrical stimulators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	02/02/2023	02/02/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	02/02/2023	02/02/2023
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	02/02/2023	02/02/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	02/02/2023	02/02/2023
					CT2985	Active	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	02/02/2023	02/02/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	02/02/2023	02/02/2023
					CT3001	Implantable metal	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	02/02/2023	02/02/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	02/02/2023	02/02/2023
					CT3002	Implantable synthetic polymer	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	02/02/2023	02/02/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	02/02/2023	02/02/2023
					CT3006	Implantable inorganic material	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	02/02/2023	02/02/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	02/02/2023	02/02/2023
					CT314	Electrically powered	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	02/02/2023	02/02/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	02/02/2023	02/02/2023
					CT321	Long-term surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	02/02/2023	02/02/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	02/02/2023	02/02/2023
					CT449	Generators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	02/02/2023	02/02/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	02/02/2023	02/02/2023
					CT981	Single use	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	02/02/2023	02/02/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	02/02/2023	02/02/2023
					CT983	Surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	02/02/2023	02/02/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	02/02/2023	02/02/2023
12545	2023	Unanticipated serious deterioration in state of health	63188	Left-ventricular wireless pacing system sensor/transmitter	CT145	Cardiovascular devices			A070801	Complete Loss of Power		
					CT1577	Wireless monitoring system transmitters			A070801	Complete Loss of Power		
					CT1684	Cardiac pacemakers and associated devices			A070801	Complete Loss of Power		
					CT2727	Active implant devices/ancillaries			A070801	Complete Loss of Power		
					CT2750	Implantable cardiac electrical stimulators and associated devices			A070801	Complete Loss of Power		
					CT2885	Physiological monitoring/measuring system transmitters/receivers/transceivers			A070801	Complete Loss of Power		
12546	2023	All other reportable incidents	47265	Dual-chamber implantable pacemaker, rate-responsive	CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	04/03/2023	04/03/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	04/03/2023	04/03/2023
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	04/03/2023	04/03/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	04/03/2023	04/03/2023
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	04/03/2023	04/03/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	04/03/2023	04/03/2023
12547	2023		47265	Dual-chamber implantable pacemaker, rate-responsive	CT141	Electrodes/Leads and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	07/03/2023	07/03/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	07/03/2023	07/03/2023
					CT145	Cardiovascular devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	07/03/2023	07/03/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	07/03/2023	07/03/2023
					CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	07/03/2023	07/03/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	07/03/2023	07/03/2023
					CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	07/03/2023	07/03/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	07/03/2023	07/03/2023
					CT233	Surgical	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	07/03/2023	07/03/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	07/03/2023	07/03/2023
					CT2406	Implantable	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	07/03/2023	07/03/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	07/03/2023	07/03/2023
					CT2407	Active implantable	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	07/03/2023	07/03/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	07/03/2023	07/03/2023
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	07/03/2023	07/03/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	07/03/2023	07/03/2023
					CT2727	Active implant devices/ancillaries	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	07/03/2023	07/03/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	07/03/2023	07/03/2023

					CT2750	Implantable cardiac electrical stimulators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	07/03/2023	07/03/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	07/03/2023	07/03/2023
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	07/03/2023	07/03/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	07/03/2023	07/03/2023
					CT2985	Active	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	07/03/2023	07/03/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	07/03/2023	07/03/2023
					CT3001	Implantable metal	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	07/03/2023	07/03/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	07/03/2023	07/03/2023
					CT3002	Implantable synthetic polymer	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	07/03/2023	07/03/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	07/03/2023	07/03/2023
					CT3006	Implantable inorganic material	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	07/03/2023	07/03/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	07/03/2023	07/03/2023
					CT314	Electrically powered	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	07/03/2023	07/03/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	07/03/2023	07/03/2023
					CT321	Long-term surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	07/03/2023	07/03/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	07/03/2023	07/03/2023
					CT449	Generators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	07/03/2023	07/03/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	07/03/2023	07/03/2023
					CT981	Single use	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	07/03/2023	07/03/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	07/03/2023	07/03/2023
					CT983	Surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	07/03/2023	07/03/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	07/03/2023	07/03/2023
12548	2023		47265	Dual-chamber implantable pacemaker, rate-responsive	CT141	Electrodes/Leads and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	25/02/2023	25/02/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	25/02/2023	25/02/2023
					CT145	Cardiovascular devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	25/02/2023	25/02/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	25/02/2023	25/02/2023
					CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	25/02/2023	25/02/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	25/02/2023	25/02/2023
					CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	25/02/2023	25/02/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	25/02/2023	25/02/2023
					CT233	Surgical	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	25/02/2023	25/02/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	25/02/2023	25/02/2023
					CT2406	Implantable	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	25/02/2023	25/02/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	25/02/2023	25/02/2023
					CT2407	Active implantable	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	25/02/2023	25/02/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	25/02/2023	25/02/2023
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	25/02/2023	25/02/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	25/02/2023	25/02/2023
					CT2727	Active implant devices/ancillaries	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	25/02/2023	25/02/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	25/02/2023	25/02/2023
					CT2750	Implantable cardiac electrical stimulators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	25/02/2023	25/02/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	25/02/2023	25/02/2023
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	25/02/2023	25/02/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	25/02/2023	25/02/2023
					CT2985	Active	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	25/02/2023	25/02/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	25/02/2023	25/02/2023
					CT3001	Implantable metal	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	25/02/2023	25/02/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	25/02/2023	25/02/2023
					CT3002	Implantable synthetic polymer	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	25/02/2023	25/02/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	25/02/2023	25/02/2023
					CT3006	Implantable inorganic material	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	25/02/2023	25/02/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	25/02/2023	25/02/2023
					CT314	Electrically powered	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	25/02/2023	25/02/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	25/02/2023	25/02/2023
					CT321	Long-term surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	25/02/2023	25/02/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	25/02/2023	25/02/2023
					CT449	Generators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	25/02/2023	25/02/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	25/02/2023	25/02/2023
					CT981	Single use	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	25/02/2023	25/02/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	25/02/2023	25/02/2023
					CT983	Surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	25/02/2023	25/02/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	25/02/2023	25/02/2023
12549	2023	All other reportable incidents	35852	Single-chamber implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
12550	2023	All other reportable incidents	47267	Single-chamber implantable pacemaker, rate-responsive	CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
12551	2023	All other reportable incidents	17882	Physiologic-monitoring defibrillation system	CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A07	Electrical /Electronic Property Problem		
							F27	Problem Identified During Non-Clinical Procedure	A07	Electrical /Electronic Property Problem		
12552	2023	All other reportable incidents	35852	Single-chamber implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E2104	Electric Shock	A070909	Oversensing		
									A071302	Inappropriate/Inadequate Shock/Stimulation		
									A090801	Signal Artifact/Noise		
							F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A070909	Oversensing		
									A071302	Inappropriate/Inadequate Shock/Stimulation		
									A090801	Signal Artifact/Noise		
							F2203	Imaging Required	A070909	Oversensing		
									A071302	Inappropriate/Inadequate Shock/Stimulation		
									A090801	Signal Artifact/Noise		
12553	2023		47265	Dual-chamber implantable pacemaker, rate-	CT141	Electrodes/Leads and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	28/02/2023	28/02/2023

				responsive			F1901	Additional Surgery	A24	No Apparent Device Problem	28/02/2023	28/02/2023
					CT145	Cardiovascular devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	28/02/2023	28/02/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	28/02/2023	28/02/2023
					CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	28/02/2023	28/02/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	28/02/2023	28/02/2023
					CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	28/02/2023	28/02/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	28/02/2023	28/02/2023
					CT233	Surgical	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	28/02/2023	28/02/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	28/02/2023	28/02/2023
					CT2406	Implantable	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	28/02/2023	28/02/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	28/02/2023	28/02/2023
					CT2407	Active implantable	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	28/02/2023	28/02/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	28/02/2023	28/02/2023
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	28/02/2023	28/02/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	28/02/2023	28/02/2023
					CT2727	Active implant devices/ancillaries	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	28/02/2023	28/02/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	28/02/2023	28/02/2023
					CT2750	Implantable cardiac electrical stimulators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	28/02/2023	28/02/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	28/02/2023	28/02/2023
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	28/02/2023	28/02/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	28/02/2023	28/02/2023
					CT2985	Active	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	28/02/2023	28/02/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	28/02/2023	28/02/2023
					CT3001	Implantable metal	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	28/02/2023	28/02/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	28/02/2023	28/02/2023
					CT3002	Implantable synthetic polymer	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	28/02/2023	28/02/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	28/02/2023	28/02/2023
					CT3006	Implantable inorganic material	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	28/02/2023	28/02/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	28/02/2023	28/02/2023
					CT314	Electrically powered	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	28/02/2023	28/02/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	28/02/2023	28/02/2023
					CT321	Long-term surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	28/02/2023	28/02/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	28/02/2023	28/02/2023
					CT449	Generators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	28/02/2023	28/02/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	28/02/2023	28/02/2023
					CT981	Single use	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	28/02/2023	28/02/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	28/02/2023	28/02/2023
					CT983	Surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	28/02/2023	28/02/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	28/02/2023	28/02/2023
12554	2023		47265	Dual-chamber implantable pacemaker, rate-responsive	CT141	Electrodes/Leads and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	25/02/2023	25/02/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	25/02/2023	25/02/2023
					CT145	Cardiovascular devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	25/02/2023	25/02/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	25/02/2023	25/02/2023
					CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	25/02/2023	25/02/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	25/02/2023	25/02/2023
					CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	25/02/2023	25/02/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	25/02/2023	25/02/2023
					CT233	Surgical	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	25/02/2023	25/02/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	25/02/2023	25/02/2023
					CT2406	Implantable	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	25/02/2023	25/02/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	25/02/2023	25/02/2023
					CT2407	Active implantable	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	25/02/2023	25/02/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	25/02/2023	25/02/2023
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	25/02/2023	25/02/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	25/02/2023	25/02/2023
					CT2727	Active implant devices/ancillaries	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	25/02/2023	25/02/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	25/02/2023	25/02/2023
					CT2750	Implantable cardiac electrical stimulators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	25/02/2023	25/02/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	25/02/2023	25/02/2023
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	25/02/2023	25/02/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	25/02/2023	25/02/2023
					CT2985	Active	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	25/02/2023	25/02/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	25/02/2023	25/02/2023
					CT3001	Implantable metal	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	25/02/2023	25/02/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	25/02/2023	25/02/2023
					CT3002	Implantable synthetic polymer	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	25/02/2023	25/02/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	25/02/2023	25/02/2023
					CT3006	Implantable inorganic material	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	25/02/2023	25/02/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	25/02/2023	25/02/2023
					CT314	Electrically powered	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	25/02/2023	25/02/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	25/02/2023	25/02/2023
					CT321	Long-term surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	25/02/2023	25/02/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	25/02/2023	25/02/2023
					CT449	Generators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	25/02/2023	25/02/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	25/02/2023	25/02/2023
					CT981	Single use	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	25/02/2023	25/02/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	25/02/2023	25/02/2023
					CT983	Surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	25/02/2023	25/02/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	25/02/2023	25/02/2023
12555	2023	All other reportable incidents	47263	Cardiac resynchronization therapy implantable pacemaker	CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A0908	Incorrect, Inadequate or Imprecise Result or Readings	30/11/2022	30/11/2022
									A160503	Inappropriate or Unexpected Reset	30/11/2022	30/11/2022
							F08	Hospitalization or Prolonged Hospitalization	A0908	Incorrect, Inadequate or Imprecise Result or Readings	30/11/2022	30/11/2022
									A160503	Inappropriate or Unexpected Reset	30/11/2022	30/11/2022
							F1905	Device Revision or Replacement	A0908	Incorrect, Inadequate or Imprecise Result or Readings	30/11/2022	30/11/2022
									A160503	Inappropriate or Unexpected Reset	30/11/2022	30/11/2022

					CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A0908	Incorrect, Inadequate or Imprecise Result or Readings	30/11/2022	30/11/2022
									A160503	Inappropriate or Unexpected Reset	30/11/2022	30/11/2022
							F08	Hospitalization or Prolonged Hospitalization	A0908	Incorrect, Inadequate or Imprecise Result or Readings	30/11/2022	30/11/2022
									A160503	Inappropriate or Unexpected Reset	30/11/2022	30/11/2022
							F1905	Device Revision or Replacement	A0908	Incorrect, Inadequate or Imprecise Result or Readings	30/11/2022	30/11/2022
									A160503	Inappropriate or Unexpected Reset	30/11/2022	30/11/2022
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A0908	Incorrect, Inadequate or Imprecise Result or Readings	30/11/2022	30/11/2022
									A160503	Inappropriate or Unexpected Reset	30/11/2022	30/11/2022
							F08	Hospitalization or Prolonged Hospitalization	A0908	Incorrect, Inadequate or Imprecise Result or Readings	30/11/2022	30/11/2022
									A160503	Inappropriate or Unexpected Reset	30/11/2022	30/11/2022
							F1905	Device Revision or Replacement	A0908	Incorrect, Inadequate or Imprecise Result or Readings	30/11/2022	30/11/2022
									A160503	Inappropriate or Unexpected Reset	30/11/2022	30/11/2022
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A0908	Incorrect, Inadequate or Imprecise Result or Readings	30/11/2022	30/11/2022
									A160503	Inappropriate or Unexpected Reset	30/11/2022	30/11/2022
							F08	Hospitalization or Prolonged Hospitalization	A0908	Incorrect, Inadequate or Imprecise Result or Readings	30/11/2022	30/11/2022
									A160503	Inappropriate or Unexpected Reset	30/11/2022	30/11/2022
							F1905	Device Revision or Replacement	A0908	Incorrect, Inadequate or Imprecise Result or Readings	30/11/2022	30/11/2022
									A160503	Inappropriate or Unexpected Reset	30/11/2022	30/11/2022
12556	2023		47265	Dual-chamber implantable pacemaker, rate-responsive	CT141	Electrodes/Leads and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	04/03/2023	04/03/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	04/03/2023	04/03/2023
					CT145	Cardiovascular devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	04/03/2023	04/03/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	04/03/2023	04/03/2023
					CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	04/03/2023	04/03/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	04/03/2023	04/03/2023
					CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	04/03/2023	04/03/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	04/03/2023	04/03/2023
					CT233	Surgical	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	04/03/2023	04/03/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	04/03/2023	04/03/2023
					CT2406	Implantable	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	04/03/2023	04/03/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	04/03/2023	04/03/2023
					CT2407	Active implantable	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	04/03/2023	04/03/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	04/03/2023	04/03/2023
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	04/03/2023	04/03/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	04/03/2023	04/03/2023
					CT2727	Active implant devices/ancillaries	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	04/03/2023	04/03/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	04/03/2023	04/03/2023
					CT2750	Implantable cardiac electrical stimulators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	04/03/2023	04/03/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	04/03/2023	04/03/2023
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	04/03/2023	04/03/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	04/03/2023	04/03/2023
					CT2985	Active	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	04/03/2023	04/03/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	04/03/2023	04/03/2023
					CT3001	Implantable metal	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	04/03/2023	04/03/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	04/03/2023	04/03/2023
					CT3002	Implantable synthetic polymer	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	04/03/2023	04/03/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	04/03/2023	04/03/2023
					CT3006	Implantable inorganic material	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	04/03/2023	04/03/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	04/03/2023	04/03/2023
					CT314	Electrically powered	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	04/03/2023	04/03/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	04/03/2023	04/03/2023
					CT321	Long-term surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	04/03/2023	04/03/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	04/03/2023	04/03/2023
					CT449	Generators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	04/03/2023	04/03/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	04/03/2023	04/03/2023
					CT981	Single use	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	04/03/2023	04/03/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	04/03/2023	04/03/2023
					CT983	Surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	04/03/2023	04/03/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	04/03/2023	04/03/2023
12557	2023		47265	Dual-chamber implantable pacemaker, rate-responsive	CT141	Electrodes/Leads and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	04/03/2023	04/03/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	04/03/2023	04/03/2023
					CT145	Cardiovascular devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	04/03/2023	04/03/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	04/03/2023	04/03/2023
					CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	04/03/2023	04/03/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	04/03/2023	04/03/2023
					CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	04/03/2023	04/03/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	04/03/2023	04/03/2023
					CT233	Surgical	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	04/03/2023	04/03/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	04/03/2023	04/03/2023
					CT2406	Implantable	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	04/03/2023	04/03/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	04/03/2023	04/03/2023
					CT2407	Active implantable	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	04/03/2023	04/03/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	04/03/2023	04/03/2023
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	04/03/2023	04/03/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	04/03/2023	04/03/2023
					CT2727	Active implant devices/ancillaries	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	04/03/2023	04/03/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	04/03/2023	04/03/2023
					CT2750	Implantable cardiac electrical stimulators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	04/03/2023	04/03/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	04/03/2023	04/03/2023
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	04/03/2023	04/03/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	04/03/2023	04/03/2023
					CT2985	Active	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	04/03/2023	04/03/2023

							F1901	Additional Surgery	A24	No Apparent Device Problem	04/03/2023	04/03/2023
					CT3001	Implantable metal	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	04/03/2023	04/03/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	04/03/2023	04/03/2023
					CT3002	Implantable synthetic polymer	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	04/03/2023	04/03/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	04/03/2023	04/03/2023
					CT3006	Implantable inorganic material	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	04/03/2023	04/03/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	04/03/2023	04/03/2023
					CT314	Electrically powered	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	04/03/2023	04/03/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	04/03/2023	04/03/2023
					CT321	Long-term surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	04/03/2023	04/03/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	04/03/2023	04/03/2023
					CT449	Generators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	04/03/2023	04/03/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	04/03/2023	04/03/2023
					CT981	Single use	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	04/03/2023	04/03/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	04/03/2023	04/03/2023
					CT983	Surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	04/03/2023	04/03/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	04/03/2023	04/03/2023
12558	2023		47265	Dual-chamber implantable pacemaker, rate-responsive	CT141	Electrodes/Leads and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	04/03/2023	04/03/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	04/03/2023	04/03/2023
					CT145	Cardiovascular devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	04/03/2023	04/03/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	04/03/2023	04/03/2023
					CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	04/03/2023	04/03/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	04/03/2023	04/03/2023
					CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	04/03/2023	04/03/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	04/03/2023	04/03/2023
					CT233	Surgical	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	04/03/2023	04/03/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	04/03/2023	04/03/2023
					CT2406	Implantable	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	04/03/2023	04/03/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	04/03/2023	04/03/2023
					CT2407	Active implantable	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	04/03/2023	04/03/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	04/03/2023	04/03/2023
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	04/03/2023	04/03/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	04/03/2023	04/03/2023
					CT2727	Active implant devices/ancillaries	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	04/03/2023	04/03/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	04/03/2023	04/03/2023
					CT2750	Implantable cardiac electrical stimulators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	04/03/2023	04/03/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	04/03/2023	04/03/2023
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	04/03/2023	04/03/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	04/03/2023	04/03/2023
					CT2985	Active	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	04/03/2023	04/03/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	04/03/2023	04/03/2023
					CT3001	Implantable metal	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	04/03/2023	04/03/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	04/03/2023	04/03/2023
					CT3002	Implantable synthetic polymer	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	04/03/2023	04/03/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	04/03/2023	04/03/2023
					CT3006	Implantable inorganic material	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	04/03/2023	04/03/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	04/03/2023	04/03/2023
					CT314	Electrically powered	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	04/03/2023	04/03/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	04/03/2023	04/03/2023
					CT321	Long-term surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	04/03/2023	04/03/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	04/03/2023	04/03/2023
					CT449	Generators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	04/03/2023	04/03/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	04/03/2023	04/03/2023
					CT981	Single use	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	04/03/2023	04/03/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	04/03/2023	04/03/2023
					CT983	Surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	04/03/2023	04/03/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	04/03/2023	04/03/2023
12559	2023		47265	Dual-chamber implantable pacemaker, rate-responsive	CT141	Electrodes/Leads and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	04/03/2023	04/03/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	04/03/2023	04/03/2023
					CT145	Cardiovascular devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	04/03/2023	04/03/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	04/03/2023	04/03/2023
					CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	04/03/2023	04/03/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	04/03/2023	04/03/2023
					CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	04/03/2023	04/03/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	04/03/2023	04/03/2023
					CT233	Surgical	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	04/03/2023	04/03/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	04/03/2023	04/03/2023
					CT2406	Implantable	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	04/03/2023	04/03/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	04/03/2023	04/03/2023
					CT2407	Active implantable	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	04/03/2023	04/03/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	04/03/2023	04/03/2023
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	04/03/2023	04/03/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	04/03/2023	04/03/2023
					CT2727	Active implant devices/ancillaries	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	04/03/2023	04/03/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	04/03/2023	04/03/2023
					CT2750	Implantable cardiac electrical stimulators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	04/03/2023	04/03/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	04/03/2023	04/03/2023
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	04/03/2023	04/03/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	04/03/2023	04/03/2023
					CT2985	Active	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	04/03/2023	04/03/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	04/03/2023	04/03/2023
					CT3001	Implantable metal	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	04/03/2023	04/03/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	04/03/2023	04/03/2023
					CT3002	Implantable synthetic polymer	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	04/03/2023	04/03/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	04/03/2023	04/03/2023
					CT3006	Implantable inorganic material	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	04/03/2023	04/03/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	04/03/2023	04/03/2023
					CT314	Electrically powered	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	04/03/2023	04/03/2023

							F1901	Additional Surgery	A24	No Apparent Device Problem	04/03/2023	04/03/2023
					CT321	Long-term surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	04/03/2023	04/03/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	04/03/2023	04/03/2023
					CT449	Generators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	04/03/2023	04/03/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	04/03/2023	04/03/2023
					CT981	Single use	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	04/03/2023	04/03/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	04/03/2023	04/03/2023
					CT983	Surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	04/03/2023	04/03/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	04/03/2023	04/03/2023
12560	2023		47265	Dual-chamber implantable pacemaker, rate-responsive	CT141	Electrodes/Leads and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	29/01/2023	29/01/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	29/01/2023	29/01/2023
					CT145	Cardiovascular devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	29/01/2023	29/01/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	29/01/2023	29/01/2023
					CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	29/01/2023	29/01/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	29/01/2023	29/01/2023
					CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	29/01/2023	29/01/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	29/01/2023	29/01/2023
					CT233	Surgical	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	29/01/2023	29/01/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	29/01/2023	29/01/2023
					CT2406	Implantable	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	29/01/2023	29/01/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	29/01/2023	29/01/2023
					CT2407	Active implantable	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	29/01/2023	29/01/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	29/01/2023	29/01/2023
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	29/01/2023	29/01/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	29/01/2023	29/01/2023
					CT2727	Active implant devices/ancillaries	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	29/01/2023	29/01/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	29/01/2023	29/01/2023
					CT2750	Implantable cardiac electrical stimulators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	29/01/2023	29/01/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	29/01/2023	29/01/2023
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	29/01/2023	29/01/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	29/01/2023	29/01/2023
					CT2985	Active	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	29/01/2023	29/01/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	29/01/2023	29/01/2023
					CT3001	Implantable metal	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	29/01/2023	29/01/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	29/01/2023	29/01/2023
					CT3002	Implantable synthetic polymer	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	29/01/2023	29/01/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	29/01/2023	29/01/2023
					CT3006	Implantable inorganic material	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	29/01/2023	29/01/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	29/01/2023	29/01/2023
					CT314	Electrically powered	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	29/01/2023	29/01/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	29/01/2023	29/01/2023
					CT321	Long-term surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	29/01/2023	29/01/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	29/01/2023	29/01/2023
					CT449	Generators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	29/01/2023	29/01/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	29/01/2023	29/01/2023
					CT981	Single use	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	29/01/2023	29/01/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	29/01/2023	29/01/2023
					CT983	Surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	29/01/2023	29/01/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	29/01/2023	29/01/2023
12561	2023	All other reportable incidents	47265	Dual-chamber implantable pacemaker, rate-responsive	CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem		
							F26	No Health Consequences or Impact	A0705	Battery Problem		
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem		
							F26	No Health Consequences or Impact	A0705	Battery Problem		
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem		
							F26	No Health Consequences or Impact	A0705	Battery Problem		
12562	2023		47265	Dual-chamber implantable pacemaker, rate-responsive	CT141	Electrodes/Leads and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem		
							F1901	Additional Surgery	A24	No Apparent Device Problem		
					CT145	Cardiovascular devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem		
							F1901	Additional Surgery	A24	No Apparent Device Problem		
					CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem		
							F1901	Additional Surgery	A24	No Apparent Device Problem		
					CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem		
							F1901	Additional Surgery	A24	No Apparent Device Problem		
					CT233	Surgical	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem		
							F1901	Additional Surgery	A24	No Apparent Device Problem		
					CT2406	Implantable	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem		
							F1901	Additional Surgery	A24	No Apparent Device Problem		
					CT2407	Active implantable	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem		
							F1901	Additional Surgery	A24	No Apparent Device Problem		
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem		
							F1901	Additional Surgery	A24	No Apparent Device Problem		
					CT2727	Active implant devices/ancillaries	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem		
							F1901	Additional Surgery	A24	No Apparent Device Problem		
					CT2750	Implantable cardiac electrical stimulators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem		
							F1901	Additional Surgery	A24	No Apparent Device Problem		
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem		
							F1901	Additional Surgery	A24	No Apparent Device Problem		
					CT2985	Active	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem		
							F1901	Additional Surgery	A24	No Apparent Device Problem		
					CT3001	Implantable metal	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem		
							F1901	Additional Surgery	A24	No Apparent Device Problem		
					CT3002	Implantable synthetic polymer	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem		
							F1901	Additional Surgery	A24	No Apparent Device Problem		
					CT3006	Implantable inorganic material	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem		
							F1901	Additional Surgery	A24	No Apparent Device Problem		
					CT314	Electrically powered	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem		
							F1901	Additional Surgery	A24	No Apparent Device Problem		
					CT321	Long-term surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem		

							F1901	Additional Surgery	A24	No Apparent Device Problem		
					CT449	Generators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem		
							F1901	Additional Surgery	A24	No Apparent Device Problem		
					CT981	Single use	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem		
							F1901	Additional Surgery	A24	No Apparent Device Problem		
					CT983	Surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem		
							F1901	Additional Surgery	A24	No Apparent Device Problem		
12563	2023	All other reportable incidents	17882	Physiologic-monitoring defibrillation system	CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A1603	Failure of Device to Self-Test		
							F27	Problem Identified During Non-Clinical Procedure	A1603	Failure of Device to Self-Test		
12564	2023	All other reportable incidents	35852	Single-chamber implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem	15/09/2022	15/09/2022
									A070504	Premature Discharge of Battery	15/09/2022	15/09/2022
							F08	Hospitalization or Prolonged Hospitalization	A0705	Battery Problem	15/09/2022	15/09/2022
									A070504	Premature Discharge of Battery	15/09/2022	15/09/2022
							F1905	Device Revision or Replacement	A0705	Battery Problem	15/09/2022	15/09/2022
									A070504	Premature Discharge of Battery	15/09/2022	15/09/2022
12565	2023	All other reportable incidents	47267	Single-chamber implantable pacemaker, rate-responsive	CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem	30/09/2022	30/09/2022
									A070504	Premature Discharge of Battery	30/09/2022	30/09/2022
							F08	Hospitalization or Prolonged Hospitalization	A0705	Battery Problem	30/09/2022	30/09/2022
									A070504	Premature Discharge of Battery	30/09/2022	30/09/2022
							F1905	Device Revision or Replacement	A0705	Battery Problem	30/09/2022	30/09/2022
									A070504	Premature Discharge of Battery	30/09/2022	30/09/2022
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem	30/09/2022	30/09/2022
									A070504	Premature Discharge of Battery	30/09/2022	30/09/2022
							F08	Hospitalization or Prolonged Hospitalization	A0705	Battery Problem	30/09/2022	30/09/2022
									A070504	Premature Discharge of Battery	30/09/2022	30/09/2022
							F1905	Device Revision or Replacement	A0705	Battery Problem	30/09/2022	30/09/2022
									A070504	Premature Discharge of Battery	30/09/2022	30/09/2022
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem	30/09/2022	30/09/2022
									A070504	Premature Discharge of Battery	30/09/2022	30/09/2022
							F08	Hospitalization or Prolonged Hospitalization	A0705	Battery Problem	30/09/2022	30/09/2022
									A070504	Premature Discharge of Battery	30/09/2022	30/09/2022
							F1905	Device Revision or Replacement	A0705	Battery Problem	30/09/2022	30/09/2022
									A070504	Premature Discharge of Battery	30/09/2022	30/09/2022
12566	2023	All other reportable incidents	47267	Single-chamber implantable pacemaker, rate-responsive	CT141	Electrodes/Leads and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem	06/09/2022	06/09/2022
									A070504	Premature Discharge of Battery	06/09/2022	06/09/2022
							F08	Hospitalization or Prolonged Hospitalization	A0705	Battery Problem	06/09/2022	06/09/2022
									A070504	Premature Discharge of Battery	06/09/2022	06/09/2022
							F1905	Device Revision or Replacement	A0705	Battery Problem	06/09/2022	06/09/2022
									A070504	Premature Discharge of Battery	06/09/2022	06/09/2022
					CT145	Cardiovascular devices	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem	06/09/2022	06/09/2022
									A070504	Premature Discharge of Battery	06/09/2022	06/09/2022
							F08	Hospitalization or Prolonged Hospitalization	A0705	Battery Problem	06/09/2022	06/09/2022
									A070504	Premature Discharge of Battery	06/09/2022	06/09/2022
							F1905	Device Revision or Replacement	A0705	Battery Problem	06/09/2022	06/09/2022
									A070504	Premature Discharge of Battery	06/09/2022	06/09/2022
					CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem	06/09/2022	06/09/2022
									A070504	Premature Discharge of Battery	06/09/2022	06/09/2022
							F08	Hospitalization or Prolonged Hospitalization	A0705	Battery Problem	06/09/2022	06/09/2022
									A070504	Premature Discharge of Battery	06/09/2022	06/09/2022
							F1905	Device Revision or Replacement	A0705	Battery Problem	06/09/2022	06/09/2022
									A070504	Premature Discharge of Battery	06/09/2022	06/09/2022
					CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem	06/09/2022	06/09/2022
									A070504	Premature Discharge of Battery	06/09/2022	06/09/2022
							F08	Hospitalization or Prolonged Hospitalization	A0705	Battery Problem	06/09/2022	06/09/2022
									A070504	Premature Discharge of Battery	06/09/2022	06/09/2022
							F1905	Device Revision or Replacement	A0705	Battery Problem	06/09/2022	06/09/2022
									A070504	Premature Discharge of Battery	06/09/2022	06/09/2022
					CT233	Surgical	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem	06/09/2022	06/09/2022
									A070504	Premature Discharge of Battery	06/09/2022	06/09/2022
							F08	Hospitalization or Prolonged Hospitalization	A0705	Battery Problem	06/09/2022	06/09/2022
									A070504	Premature Discharge of Battery	06/09/2022	06/09/2022
							F1905	Device Revision or Replacement	A0705	Battery Problem	06/09/2022	06/09/2022
									A070504	Premature Discharge of Battery	06/09/2022	06/09/2022
					CT2406	Implantable	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem	06/09/2022	06/09/2022
									A070504	Premature Discharge of Battery	06/09/2022	06/09/2022
							F08	Hospitalization or Prolonged Hospitalization	A0705	Battery Problem	06/09/2022	06/09/2022
									A070504	Premature Discharge of Battery	06/09/2022	06/09/2022
							F1905	Device Revision or Replacement	A0705	Battery Problem	06/09/2022	06/09/2022
									A070504	Premature Discharge of Battery	06/09/2022	06/09/2022
					CT2407	Active implantable	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem	06/09/2022	06/09/2022
									A070504	Premature Discharge of Battery	06/09/2022	06/09/2022
							F08	Hospitalization or Prolonged Hospitalization	A0705	Battery Problem	06/09/2022	06/09/2022
									A070504	Premature Discharge of Battery	06/09/2022	06/09/2022
							F1905	Device Revision or Replacement	A0705	Battery Problem	06/09/2022	06/09/2022
									A070504	Premature Discharge of Battery	06/09/2022	06/09/2022
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem	06/09/2022	06/09/2022
									A070504	Premature Discharge of Battery	06/09/2022	06/09/2022
							F08	Hospitalization or Prolonged Hospitalization	A0705	Battery Problem	06/09/2022	06/09/2022
									A070504	Premature Discharge of Battery	06/09/2022	06/09/2022
							F1905	Device Revision or Replacement	A0705	Battery Problem	06/09/2022	06/09/2022
									A070504	Premature Discharge of Battery	06/09/2022	06/09/2022
					CT2727	Active implant devices/ancillaries	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem	06/09/2022	06/09/2022
									A070504	Premature Discharge of Battery	06/09/2022	06/09/2022
							F08	Hospitalization or Prolonged Hospitalization	A0705	Battery Problem	06/09/2022	06/09/2022
									A070504	Premature Discharge of Battery	06/09/2022	06/09/2022
							F1905	Device Revision or Replacement	A0705	Battery Problem	06/09/2022	06/09/2022
									A070504	Premature Discharge of Battery	06/09/2022	06/09/2022
					CT2750	Implantable cardiac electrical stimulators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem	06/09/2022	06/09/2022

								A070504	Premature Discharge of Battery	06/09/2022	06/09/2022
						F08	Hospitalization or Prolonged Hospitalization	A0705	Battery Problem	06/09/2022	06/09/2022
								A070504	Premature Discharge of Battery	06/09/2022	06/09/2022
						F1905	Device Revision or Replacement	A0705	Battery Problem	06/09/2022	06/09/2022
								A070504	Premature Discharge of Battery	06/09/2022	06/09/2022
				CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem	06/09/2022	06/09/2022
								A070504	Premature Discharge of Battery	06/09/2022	06/09/2022
						F08	Hospitalization or Prolonged Hospitalization	A0705	Battery Problem	06/09/2022	06/09/2022
								A070504	Premature Discharge of Battery	06/09/2022	06/09/2022
						F1905	Device Revision or Replacement	A0705	Battery Problem	06/09/2022	06/09/2022
								A070504	Premature Discharge of Battery	06/09/2022	06/09/2022
				CT2985	Active	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem	06/09/2022	06/09/2022
								A070504	Premature Discharge of Battery	06/09/2022	06/09/2022
						F08	Hospitalization or Prolonged Hospitalization	A0705	Battery Problem	06/09/2022	06/09/2022
								A070504	Premature Discharge of Battery	06/09/2022	06/09/2022
						F1905	Device Revision or Replacement	A0705	Battery Problem	06/09/2022	06/09/2022
								A070504	Premature Discharge of Battery	06/09/2022	06/09/2022
				CT3001	Implantable metal	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem	06/09/2022	06/09/2022
								A070504	Premature Discharge of Battery	06/09/2022	06/09/2022
						F08	Hospitalization or Prolonged Hospitalization	A0705	Battery Problem	06/09/2022	06/09/2022
								A070504	Premature Discharge of Battery	06/09/2022	06/09/2022
						F1905	Device Revision or Replacement	A0705	Battery Problem	06/09/2022	06/09/2022
								A070504	Premature Discharge of Battery	06/09/2022	06/09/2022
				CT3002	Implantable synthetic polymer	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem	06/09/2022	06/09/2022
								A070504	Premature Discharge of Battery	06/09/2022	06/09/2022
						F08	Hospitalization or Prolonged Hospitalization	A0705	Battery Problem	06/09/2022	06/09/2022
								A070504	Premature Discharge of Battery	06/09/2022	06/09/2022
						F1905	Device Revision or Replacement	A0705	Battery Problem	06/09/2022	06/09/2022
								A070504	Premature Discharge of Battery	06/09/2022	06/09/2022
				CT3006	Implantable inorganic material	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem	06/09/2022	06/09/2022
								A070504	Premature Discharge of Battery	06/09/2022	06/09/2022
						F08	Hospitalization or Prolonged Hospitalization	A0705	Battery Problem	06/09/2022	06/09/2022
								A070504	Premature Discharge of Battery	06/09/2022	06/09/2022
						F1905	Device Revision or Replacement	A0705	Battery Problem	06/09/2022	06/09/2022
								A070504	Premature Discharge of Battery	06/09/2022	06/09/2022
				CT314	Electrically powered	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem	06/09/2022	06/09/2022
								A070504	Premature Discharge of Battery	06/09/2022	06/09/2022
						F08	Hospitalization or Prolonged Hospitalization	A0705	Battery Problem	06/09/2022	06/09/2022
								A070504	Premature Discharge of Battery	06/09/2022	06/09/2022
						F1905	Device Revision or Replacement	A0705	Battery Problem	06/09/2022	06/09/2022
								A070504	Premature Discharge of Battery	06/09/2022	06/09/2022
				CT321	Long-term surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem	06/09/2022	06/09/2022
								A070504	Premature Discharge of Battery	06/09/2022	06/09/2022
						F08	Hospitalization or Prolonged Hospitalization	A0705	Battery Problem	06/09/2022	06/09/2022
								A070504	Premature Discharge of Battery	06/09/2022	06/09/2022
						F1905	Device Revision or Replacement	A0705	Battery Problem	06/09/2022	06/09/2022
								A070504	Premature Discharge of Battery	06/09/2022	06/09/2022
				CT449	Generators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem	06/09/2022	06/09/2022
								A070504	Premature Discharge of Battery	06/09/2022	06/09/2022
						F08	Hospitalization or Prolonged Hospitalization	A0705	Battery Problem	06/09/2022	06/09/2022
								A070504	Premature Discharge of Battery	06/09/2022	06/09/2022
						F1905	Device Revision or Replacement	A0705	Battery Problem	06/09/2022	06/09/2022
								A070504	Premature Discharge of Battery	06/09/2022	06/09/2022
				CT981	Single use	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem	06/09/2022	06/09/2022
								A070504	Premature Discharge of Battery	06/09/2022	06/09/2022
						F08	Hospitalization or Prolonged Hospitalization	A0705	Battery Problem	06/09/2022	06/09/2022
								A070504	Premature Discharge of Battery	06/09/2022	06/09/2022
						F1905	Device Revision or Replacement	A0705	Battery Problem	06/09/2022	06/09/2022
								A070504	Premature Discharge of Battery	06/09/2022	06/09/2022
				CT983	Surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem	06/09/2022	06/09/2022
								A070504	Premature Discharge of Battery	06/09/2022	06/09/2022
						F08	Hospitalization or Prolonged Hospitalization	A0705	Battery Problem	06/09/2022	06/09/2022
								A070504	Premature Discharge of Battery	06/09/2022	06/09/2022
						F1905	Device Revision or Replacement	A0705	Battery Problem	06/09/2022	06/09/2022
								A070504	Premature Discharge of Battery	06/09/2022	06/09/2022
12567	2023	All other reportable incidents	47267	Single-chamber implantable pacemaker, rate-responsive	CT141	Electrodes/Leads and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem	19/06/2023	19/06/2023
								A070504	Premature Discharge of Battery	19/06/2023	19/06/2023
						F08	Hospitalization or Prolonged Hospitalization	A0705	Battery Problem	19/06/2023	19/06/2023
								A070504	Premature Discharge of Battery	19/06/2023	19/06/2023
						F1905	Device Revision or Replacement	A0705	Battery Problem	19/06/2023	19/06/2023
								A070504	Premature Discharge of Battery	19/06/2023	19/06/2023
				CT145	Cardiovascular devices	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem	19/06/2023	19/06/2023
								A070504	Premature Discharge of Battery	19/06/2023	19/06/2023
						F08	Hospitalization or Prolonged Hospitalization	A0705	Battery Problem	19/06/2023	19/06/2023
								A070504	Premature Discharge of Battery	19/06/2023	19/06/2023
						F1905	Device Revision or Replacement	A0705	Battery Problem	19/06/2023	19/06/2023
								A070504	Premature Discharge of Battery	19/06/2023	19/06/2023
				CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem	19/06/2023	19/06/2023
								A070504	Premature Discharge of Battery	19/06/2023	19/06/2023
						F08	Hospitalization or Prolonged Hospitalization	A0705	Battery Problem	19/06/2023	19/06/2023
								A070504	Premature Discharge of Battery	19/06/2023	19/06/2023
						F1905	Device Revision or Replacement	A0705	Battery Problem	19/06/2023	19/06/2023
								A070504	Premature Discharge of Battery	19/06/2023	19/06/2023
				CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem	19/06/2023	19/06/2023
								A070504	Premature Discharge of Battery	19/06/2023	19/06/2023
						F08	Hospitalization or Prolonged Hospitalization	A0705	Battery Problem	19/06/2023	19/06/2023
								A070504	Premature Discharge of Battery	19/06/2023	19/06/2023
						F1905	Device Revision or Replacement	A0705	Battery Problem	19/06/2023	19/06/2023

								A070504	Premature Discharge of Battery	19/06/2023	19/06/2023
					CT233	Surgical	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem	19/06/2023	19/06/2023
									A070504	Premature Discharge of Battery	19/06/2023	19/06/2023
							F08	Hospitalization or Prolonged Hospitalization	A0705	Battery Problem	19/06/2023	19/06/2023
									A070504	Premature Discharge of Battery	19/06/2023	19/06/2023
							F1905	Device Revision or Replacement	A0705	Battery Problem	19/06/2023	19/06/2023
									A070504	Premature Discharge of Battery	19/06/2023	19/06/2023
					CT2406	Implantable	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem	19/06/2023	19/06/2023
									A070504	Premature Discharge of Battery	19/06/2023	19/06/2023
							F08	Hospitalization or Prolonged Hospitalization	A0705	Battery Problem	19/06/2023	19/06/2023
									A070504	Premature Discharge of Battery	19/06/2023	19/06/2023
							F1905	Device Revision or Replacement	A0705	Battery Problem	19/06/2023	19/06/2023
									A070504	Premature Discharge of Battery	19/06/2023	19/06/2023
					CT2407	Active implantable	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem	19/06/2023	19/06/2023
									A070504	Premature Discharge of Battery	19/06/2023	19/06/2023
							F08	Hospitalization or Prolonged Hospitalization	A0705	Battery Problem	19/06/2023	19/06/2023
									A070504	Premature Discharge of Battery	19/06/2023	19/06/2023
							F1905	Device Revision or Replacement	A0705	Battery Problem	19/06/2023	19/06/2023
									A070504	Premature Discharge of Battery	19/06/2023	19/06/2023
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem	19/06/2023	19/06/2023
									A070504	Premature Discharge of Battery	19/06/2023	19/06/2023
							F08	Hospitalization or Prolonged Hospitalization	A0705	Battery Problem	19/06/2023	19/06/2023
									A070504	Premature Discharge of Battery	19/06/2023	19/06/2023
							F1905	Device Revision or Replacement	A0705	Battery Problem	19/06/2023	19/06/2023
									A070504	Premature Discharge of Battery	19/06/2023	19/06/2023
					CT2727	Active implant devices/ancillaries	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem	19/06/2023	19/06/2023
									A070504	Premature Discharge of Battery	19/06/2023	19/06/2023
							F08	Hospitalization or Prolonged Hospitalization	A0705	Battery Problem	19/06/2023	19/06/2023
									A070504	Premature Discharge of Battery	19/06/2023	19/06/2023
							F1905	Device Revision or Replacement	A0705	Battery Problem	19/06/2023	19/06/2023
									A070504	Premature Discharge of Battery	19/06/2023	19/06/2023
					CT2750	Implantable cardiac electrical stimulators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem	19/06/2023	19/06/2023
									A070504	Premature Discharge of Battery	19/06/2023	19/06/2023
							F08	Hospitalization or Prolonged Hospitalization	A0705	Battery Problem	19/06/2023	19/06/2023
									A070504	Premature Discharge of Battery	19/06/2023	19/06/2023
							F1905	Device Revision or Replacement	A0705	Battery Problem	19/06/2023	19/06/2023
									A070504	Premature Discharge of Battery	19/06/2023	19/06/2023
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem	19/06/2023	19/06/2023
									A070504	Premature Discharge of Battery	19/06/2023	19/06/2023
							F08	Hospitalization or Prolonged Hospitalization	A0705	Battery Problem	19/06/2023	19/06/2023
									A070504	Premature Discharge of Battery	19/06/2023	19/06/2023
							F1905	Device Revision or Replacement	A0705	Battery Problem	19/06/2023	19/06/2023
									A070504	Premature Discharge of Battery	19/06/2023	19/06/2023
					CT2985	Active	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem	19/06/2023	19/06/2023
									A070504	Premature Discharge of Battery	19/06/2023	19/06/2023
							F08	Hospitalization or Prolonged Hospitalization	A0705	Battery Problem	19/06/2023	19/06/2023
									A070504	Premature Discharge of Battery	19/06/2023	19/06/2023
							F1905	Device Revision or Replacement	A0705	Battery Problem	19/06/2023	19/06/2023
									A070504	Premature Discharge of Battery	19/06/2023	19/06/2023
					CT3001	Implantable metal	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem	19/06/2023	19/06/2023
									A070504	Premature Discharge of Battery	19/06/2023	19/06/2023
							F08	Hospitalization or Prolonged Hospitalization	A0705	Battery Problem	19/06/2023	19/06/2023
									A070504	Premature Discharge of Battery	19/06/2023	19/06/2023
							F1905	Device Revision or Replacement	A0705	Battery Problem	19/06/2023	19/06/2023
									A070504	Premature Discharge of Battery	19/06/2023	19/06/2023
					CT3002	Implantable synthetic polymer	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem	19/06/2023	19/06/2023
									A070504	Premature Discharge of Battery	19/06/2023	19/06/2023
							F08	Hospitalization or Prolonged Hospitalization	A0705	Battery Problem	19/06/2023	19/06/2023
									A070504	Premature Discharge of Battery	19/06/2023	19/06/2023
							F1905	Device Revision or Replacement	A0705	Battery Problem	19/06/2023	19/06/2023
									A070504	Premature Discharge of Battery	19/06/2023	19/06/2023
					CT3006	Implantable inorganic material	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem	19/06/2023	19/06/2023
									A070504	Premature Discharge of Battery	19/06/2023	19/06/2023
							F08	Hospitalization or Prolonged Hospitalization	A0705	Battery Problem	19/06/2023	19/06/2023
									A070504	Premature Discharge of Battery	19/06/2023	19/06/2023
							F1905	Device Revision or Replacement	A0705	Battery Problem	19/06/2023	19/06/2023
									A070504	Premature Discharge of Battery	19/06/2023	19/06/2023
					CT314	Electrically powered	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem	19/06/2023	19/06/2023
									A070504	Premature Discharge of Battery	19/06/2023	19/06/2023
							F08	Hospitalization or Prolonged Hospitalization	A0705	Battery Problem	19/06/2023	19/06/2023
									A070504	Premature Discharge of Battery	19/06/2023	19/06/2023
							F1905	Device Revision or Replacement	A0705	Battery Problem	19/06/2023	19/06/2023
									A070504	Premature Discharge of Battery	19/06/2023	19/06/2023
					CT321	Long-term surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem	19/06/2023	19/06/2023
									A070504	Premature Discharge of Battery	19/06/2023	19/06/2023
							F08	Hospitalization or Prolonged Hospitalization	A0705	Battery Problem	19/06/2023	19/06/2023
									A070504	Premature Discharge of Battery	19/06/2023	19/06/2023
							F1905	Device Revision or Replacement	A0705	Battery Problem	19/06/2023	19/06/2023
									A070504	Premature Discharge of Battery	19/06/2023	19/06/2023
					CT449	Generators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem	19/06/2023	19/06/2023
									A070504	Premature Discharge of Battery	19/06/2023	19/06/2023
							F08	Hospitalization or Prolonged Hospitalization	A0705	Battery Problem	19/06/2023	19/06/2023
									A070504	Premature Discharge of Battery	19/06/2023	19/06/2023
							F1905	Device Revision or Replacement	A0705	Battery Problem	19/06/2023	19/06/2023
									A070504	Premature Discharge of Battery	19/06/2023	19/06/2023
					CT981	Single use	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem	19/06/2023	19/06/2023
									A070504	Premature Discharge of Battery	19/06/2023	19/06/2023
							F08	Hospitalization or Prolonged Hospitalization	A0705	Battery Problem	19/06/2023	19/06/2023

								A070504	Premature Discharge of Battery	19/06/2023	19/06/2023
						F1905	Device Revision or Replacement	A0705	Battery Problem	19/06/2023	19/06/2023
								A070504	Premature Discharge of Battery	19/06/2023	19/06/2023
				CT983	Surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem	19/06/2023	19/06/2023
								A070504	Premature Discharge of Battery	19/06/2023	19/06/2023
						F08	Hospitalization or Prolonged Hospitalization	A0705	Battery Problem	19/06/2023	19/06/2023
								A070504	Premature Discharge of Battery	19/06/2023	19/06/2023
						F1905	Device Revision or Replacement	A0705	Battery Problem	19/06/2023	19/06/2023
								A070504	Premature Discharge of Battery	19/06/2023	19/06/2023
12568	2023		47265	Dual-chamber implantable pacemaker, rate-responsive	CT141	Electrodes/Leads and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	27/02/2023	27/02/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	27/02/2023	27/02/2023
					CT145	Cardiovascular devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	27/02/2023	27/02/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	27/02/2023	27/02/2023
					CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	27/02/2023	27/02/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	27/02/2023	27/02/2023
					CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	27/02/2023	27/02/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	27/02/2023	27/02/2023
					CT233	Surgical	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	27/02/2023	27/02/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	27/02/2023	27/02/2023
					CT2406	Implantable	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	27/02/2023	27/02/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	27/02/2023	27/02/2023
					CT2407	Active implantable	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	27/02/2023	27/02/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	27/02/2023	27/02/2023
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	27/02/2023	27/02/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	27/02/2023	27/02/2023
					CT2727	Active implant devices/ancillaries	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	27/02/2023	27/02/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	27/02/2023	27/02/2023
					CT2750	Implantable cardiac electrical stimulators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	27/02/2023	27/02/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	27/02/2023	27/02/2023
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	27/02/2023	27/02/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	27/02/2023	27/02/2023
					CT2985	Active	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	27/02/2023	27/02/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	27/02/2023	27/02/2023
					CT3001	Implantable metal	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	27/02/2023	27/02/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	27/02/2023	27/02/2023
					CT3002	Implantable synthetic polymer	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	27/02/2023	27/02/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	27/02/2023	27/02/2023
					CT3006	Implantable inorganic material	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	27/02/2023	27/02/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	27/02/2023	27/02/2023
					CT314	Electrically powered	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	27/02/2023	27/02/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	27/02/2023	27/02/2023
					CT321	Long-term surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	27/02/2023	27/02/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	27/02/2023	27/02/2023
					CT449	Generators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	27/02/2023	27/02/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	27/02/2023	27/02/2023
					CT981	Single use	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	27/02/2023	27/02/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	27/02/2023	27/02/2023
					CT983	Surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	27/02/2023	27/02/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	27/02/2023	27/02/2023
12569	2023		35852	Single-chamber implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A072201	High Impedance		
							F1901	Additional Surgery	A072201	High Impedance		
12570	2023	All other reportable incidents	35852	Single-chamber implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A0508	Noise, Audible		
									A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A0508	Noise, Audible		
									A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
12571	2023		47267	Single-chamber implantable pacemaker, rate-responsive	CT141	Electrodes/Leads and associated devices			A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
					CT145	Cardiovascular devices			A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
					CT1684	Cardiac pacemakers and associated devices			A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
					CT1685	Cardiac pacemakers			A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
					CT233	Surgical			A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
					CT2406	Implantable			A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
					CT2407	Active implantable			A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
					CT257	Implantable cardiac pacemakers			A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
					CT2727	Active implant devices/ancillaries			A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
					CT2750	Implantable cardiac electrical stimulators and associated devices			A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
					CT2866	Implantable cardiac electrical stimulators			A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
					CT2985	Active			A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
					CT3001	Implantable metal			A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
					CT3002	Implantable synthetic polymer			A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
					CT3006	Implantable inorganic material			A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
					CT314	Electrically powered			A0705	Battery Problem		

									A070504	Premature Discharge of Battery		
				CT321	Long-term surgical invasive				A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
				CT449	Generators and associated devices				A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
				CT981	Single use				A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
				CT983	Surgical invasive				A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
12572	2023		47265	Dual-chamber implantable pacemaker, rate-responsive	CT141	Electrodes/Leads and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	03/02/2023	03/02/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	03/02/2023	03/02/2023
					CT145	Cardiovascular devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	03/02/2023	03/02/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	03/02/2023	03/02/2023
					CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	03/02/2023	03/02/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	03/02/2023	03/02/2023
					CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	03/02/2023	03/02/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	03/02/2023	03/02/2023
					CT233	Surgical	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	03/02/2023	03/02/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	03/02/2023	03/02/2023
					CT2406	Implantable	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	03/02/2023	03/02/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	03/02/2023	03/02/2023
					CT2407	Active implantable	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	03/02/2023	03/02/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	03/02/2023	03/02/2023
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	03/02/2023	03/02/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	03/02/2023	03/02/2023
					CT2727	Active implant devices/ancillaries	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	03/02/2023	03/02/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	03/02/2023	03/02/2023
					CT2750	Implantable cardiac electrical stimulators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	03/02/2023	03/02/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	03/02/2023	03/02/2023
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	03/02/2023	03/02/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	03/02/2023	03/02/2023
					CT2985	Active	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	03/02/2023	03/02/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	03/02/2023	03/02/2023
					CT3001	Implantable metal	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	03/02/2023	03/02/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	03/02/2023	03/02/2023
					CT3002	Implantable synthetic polymer	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	03/02/2023	03/02/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	03/02/2023	03/02/2023
					CT3006	Implantable inorganic material	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	03/02/2023	03/02/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	03/02/2023	03/02/2023
					CT314	Electrically powered	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	03/02/2023	03/02/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	03/02/2023	03/02/2023
					CT321	Long-term surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	03/02/2023	03/02/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	03/02/2023	03/02/2023
					CT449	Generators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	03/02/2023	03/02/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	03/02/2023	03/02/2023
					CT981	Single use	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	03/02/2023	03/02/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	03/02/2023	03/02/2023
					CT983	Surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	03/02/2023	03/02/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	03/02/2023	03/02/2023
12573	2023		47265	Dual-chamber implantable pacemaker, rate-responsive	CT141	Electrodes/Leads and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	27/01/2023	27/01/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	27/01/2023	27/01/2023
					CT145	Cardiovascular devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	27/01/2023	27/01/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	27/01/2023	27/01/2023
					CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	27/01/2023	27/01/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	27/01/2023	27/01/2023
					CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	27/01/2023	27/01/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	27/01/2023	27/01/2023
					CT233	Surgical	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	27/01/2023	27/01/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	27/01/2023	27/01/2023
					CT2406	Implantable	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	27/01/2023	27/01/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	27/01/2023	27/01/2023
					CT2407	Active implantable	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	27/01/2023	27/01/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	27/01/2023	27/01/2023
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	27/01/2023	27/01/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	27/01/2023	27/01/2023
					CT2727	Active implant devices/ancillaries	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	27/01/2023	27/01/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	27/01/2023	27/01/2023
					CT2750	Implantable cardiac electrical stimulators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	27/01/2023	27/01/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	27/01/2023	27/01/2023
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	27/01/2023	27/01/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	27/01/2023	27/01/2023
					CT2985	Active	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	27/01/2023	27/01/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	27/01/2023	27/01/2023
					CT3001	Implantable metal	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	27/01/2023	27/01/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	27/01/2023	27/01/2023
					CT3002	Implantable synthetic polymer	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	27/01/2023	27/01/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	27/01/2023	27/01/2023
					CT3006	Implantable inorganic material	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	27/01/2023	27/01/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	27/01/2023	27/01/2023
					CT314	Electrically powered	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	27/01/2023	27/01/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	27/01/2023	27/01/2023
					CT321	Long-term surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	27/01/2023	27/01/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	27/01/2023	27/01/2023
					CT449	Generators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	27/01/2023	27/01/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	27/01/2023	27/01/2023
					CT981	Single use	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	27/01/2023	27/01/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	27/01/2023	27/01/2023
					CT983	Surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	27/01/2023	27/01/2023

							F1901	Additional Surgery	A24	No Apparent Device Problem	27/01/2023	27/01/2023
12574	2023	All other reportable incidents	35852	Single-chamber implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	21/11/2022	21/11/2022
									A0908	Incorrect, Inadequate or Imprecise Result or Readings	21/11/2022	21/11/2022
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	21/11/2022	21/11/2022
									A0908	Incorrect, Inadequate or Imprecise Result or Readings	21/11/2022	21/11/2022
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	21/11/2022	21/11/2022
									A0908	Incorrect, Inadequate or Imprecise Result or Readings	21/11/2022	21/11/2022
12575	2023		47265	Dual-chamber implantable pacemaker, rate-responsive	CT141	Electrodes/Leads and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	27/02/2023	27/02/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	27/02/2023	27/02/2023
					CT145	Cardiovascular devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	27/02/2023	27/02/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	27/02/2023	27/02/2023
					CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	27/02/2023	27/02/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	27/02/2023	27/02/2023
					CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	27/02/2023	27/02/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	27/02/2023	27/02/2023
					CT233	Surgical	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	27/02/2023	27/02/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	27/02/2023	27/02/2023
					CT2406	Implantable	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	27/02/2023	27/02/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	27/02/2023	27/02/2023
					CT2407	Active implantable	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	27/02/2023	27/02/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	27/02/2023	27/02/2023
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	27/02/2023	27/02/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	27/02/2023	27/02/2023
					CT2727	Active implant devices/ancillaries	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	27/02/2023	27/02/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	27/02/2023	27/02/2023
					CT2750	Implantable cardiac electrical stimulators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	27/02/2023	27/02/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	27/02/2023	27/02/2023
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	27/02/2023	27/02/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	27/02/2023	27/02/2023
					CT2985	Active	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	27/02/2023	27/02/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	27/02/2023	27/02/2023
					CT3001	Implantable metal	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	27/02/2023	27/02/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	27/02/2023	27/02/2023
					CT3002	Implantable synthetic polymer	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	27/02/2023	27/02/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	27/02/2023	27/02/2023
					CT3006	Implantable inorganic material	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	27/02/2023	27/02/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	27/02/2023	27/02/2023
					CT314	Electrically powered	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	27/02/2023	27/02/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	27/02/2023	27/02/2023
					CT321	Long-term surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	27/02/2023	27/02/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	27/02/2023	27/02/2023
					CT449	Generators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	27/02/2023	27/02/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	27/02/2023	27/02/2023
					CT981	Single use	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	27/02/2023	27/02/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	27/02/2023	27/02/2023
					CT983	Surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	27/02/2023	27/02/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	27/02/2023	27/02/2023
12576	2023	All other reportable incidents	47265	Dual-chamber implantable pacemaker, rate-responsive	CT141	Electrodes/Leads and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	23/02/2023	23/02/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	23/02/2023	23/02/2023
					CT145	Cardiovascular devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	23/02/2023	23/02/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	23/02/2023	23/02/2023
					CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	23/02/2023	23/02/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	23/02/2023	23/02/2023
					CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	23/02/2023	23/02/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	23/02/2023	23/02/2023
					CT233	Surgical	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	23/02/2023	23/02/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	23/02/2023	23/02/2023
					CT2406	Implantable	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	23/02/2023	23/02/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	23/02/2023	23/02/2023
					CT2407	Active implantable	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	23/02/2023	23/02/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	23/02/2023	23/02/2023
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	23/02/2023	23/02/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	23/02/2023	23/02/2023
					CT2727	Active implant devices/ancillaries	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	23/02/2023	23/02/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	23/02/2023	23/02/2023
					CT2750	Implantable cardiac electrical stimulators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	23/02/2023	23/02/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	23/02/2023	23/02/2023
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	23/02/2023	23/02/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	23/02/2023	23/02/2023
					CT2985	Active	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	23/02/2023	23/02/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	23/02/2023	23/02/2023
					CT3001	Implantable metal	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	23/02/2023	23/02/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	23/02/2023	23/02/2023
					CT3002	Implantable synthetic polymer	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	23/02/2023	23/02/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	23/02/2023	23/02/2023
					CT3006	Implantable inorganic material	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	23/02/2023	23/02/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	23/02/2023	23/02/2023
					CT314	Electrically powered	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	23/02/2023	23/02/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	23/02/2023	23/02/2023
					CT321	Long-term surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	23/02/2023	23/02/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	23/02/2023	23/02/2023
					CT449	Generators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	23/02/2023	23/02/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	23/02/2023	23/02/2023
					CT981	Single use	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	23/02/2023	23/02/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	23/02/2023	23/02/2023

					CT983	Surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	23/02/2023	23/02/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	23/02/2023	23/02/2023
12577	2023		47265	Dual-chamber implantable pacemaker, rate-responsive	CT141	Electrodes/Leads and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	29/01/2023	29/01/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	29/01/2023	29/01/2023
					CT145	Cardiovascular devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	29/01/2023	29/01/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	29/01/2023	29/01/2023
					CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	29/01/2023	29/01/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	29/01/2023	29/01/2023
					CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	29/01/2023	29/01/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	29/01/2023	29/01/2023
					CT233	Surgical	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	29/01/2023	29/01/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	29/01/2023	29/01/2023
					CT2406	Implantable	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	29/01/2023	29/01/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	29/01/2023	29/01/2023
					CT2407	Active implantable	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	29/01/2023	29/01/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	29/01/2023	29/01/2023
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	29/01/2023	29/01/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	29/01/2023	29/01/2023
					CT2727	Active implant devices/ancillaries	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	29/01/2023	29/01/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	29/01/2023	29/01/2023
					CT2750	Implantable cardiac electrical stimulators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	29/01/2023	29/01/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	29/01/2023	29/01/2023
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	29/01/2023	29/01/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	29/01/2023	29/01/2023
					CT2985	Active	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	29/01/2023	29/01/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	29/01/2023	29/01/2023
					CT3001	Implantable metal	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	29/01/2023	29/01/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	29/01/2023	29/01/2023
					CT3002	Implantable synthetic polymer	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	29/01/2023	29/01/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	29/01/2023	29/01/2023
					CT3006	Implantable inorganic material	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	29/01/2023	29/01/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	29/01/2023	29/01/2023
					CT314	Electrically powered	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	29/01/2023	29/01/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	29/01/2023	29/01/2023
					CT321	Long-term surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	29/01/2023	29/01/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	29/01/2023	29/01/2023
					CT449	Generators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	29/01/2023	29/01/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	29/01/2023	29/01/2023
					CT981	Single use	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	29/01/2023	29/01/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	29/01/2023	29/01/2023
					CT983	Surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	29/01/2023	29/01/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	29/01/2023	29/01/2023
12578	2023	All other reportable incidents	47265	Dual-chamber implantable pacemaker, rate-responsive	CT141	Electrodes/Leads and associated devices	E0601	Arrhythmia	A0712	Pacing Problem		
							F08	Hospitalization or Prolonged Hospitalization	A0712	Pacing Problem		
					CT145	Cardiovascular devices	E0601	Arrhythmia	A0712	Pacing Problem		
							F08	Hospitalization or Prolonged Hospitalization	A0712	Pacing Problem		
					CT1684	Cardiac pacemakers and associated devices	E0601	Arrhythmia	A0712	Pacing Problem		
							F08	Hospitalization or Prolonged Hospitalization	A0712	Pacing Problem		
					CT1685	Cardiac pacemakers	E0601	Arrhythmia	A0712	Pacing Problem		
							F08	Hospitalization or Prolonged Hospitalization	A0712	Pacing Problem		
					CT233	Surgical	E0601	Arrhythmia	A0712	Pacing Problem		
							F08	Hospitalization or Prolonged Hospitalization	A0712	Pacing Problem		
					CT2406	Implantable	E0601	Arrhythmia	A0712	Pacing Problem		
							F08	Hospitalization or Prolonged Hospitalization	A0712	Pacing Problem		
					CT2407	Active implantable	E0601	Arrhythmia	A0712	Pacing Problem		
							F08	Hospitalization or Prolonged Hospitalization	A0712	Pacing Problem		
					CT257	Implantable cardiac pacemakers	E0601	Arrhythmia	A0712	Pacing Problem		
							F08	Hospitalization or Prolonged Hospitalization	A0712	Pacing Problem		
					CT2727	Active implant devices/ancillaries	E0601	Arrhythmia	A0712	Pacing Problem		
							F08	Hospitalization or Prolonged Hospitalization	A0712	Pacing Problem		
					CT2750	Implantable cardiac electrical stimulators and associated devices	E0601	Arrhythmia	A0712	Pacing Problem		
							F08	Hospitalization or Prolonged Hospitalization	A0712	Pacing Problem		
					CT2866	Implantable cardiac electrical stimulators	E0601	Arrhythmia	A0712	Pacing Problem		
							F08	Hospitalization or Prolonged Hospitalization	A0712	Pacing Problem		
					CT2985	Active	E0601	Arrhythmia	A0712	Pacing Problem		
							F08	Hospitalization or Prolonged Hospitalization	A0712	Pacing Problem		
					CT3001	Implantable metal	E0601	Arrhythmia	A0712	Pacing Problem		
							F08	Hospitalization or Prolonged Hospitalization	A0712	Pacing Problem		
					CT3002	Implantable synthetic polymer	E0601	Arrhythmia	A0712	Pacing Problem		
							F08	Hospitalization or Prolonged Hospitalization	A0712	Pacing Problem		
					CT3006	Implantable inorganic material	E0601	Arrhythmia	A0712	Pacing Problem		
							F08	Hospitalization or Prolonged Hospitalization	A0712	Pacing Problem		
					CT314	Electrically powered	E0601	Arrhythmia	A0712	Pacing Problem		
							F08	Hospitalization or Prolonged Hospitalization	A0712	Pacing Problem		
					CT321	Long-term surgical invasive	E0601	Arrhythmia	A0712	Pacing Problem		
							F08	Hospitalization or Prolonged Hospitalization	A0712	Pacing Problem		
					CT449	Generators and associated devices	E0601	Arrhythmia	A0712	Pacing Problem		
							F08	Hospitalization or Prolonged Hospitalization	A0712	Pacing Problem		
					CT981	Single use	E0601	Arrhythmia	A0712	Pacing Problem		
							F08	Hospitalization or Prolonged Hospitalization	A0712	Pacing Problem		
					CT983	Surgical invasive	E0601	Arrhythmia	A0712	Pacing Problem		
							F08	Hospitalization or Prolonged Hospitalization	A0712	Pacing Problem		
12579	2023	All other reportable incidents	47265	Dual-chamber implantable pacemaker, rate-responsive	CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		

12580	2023	All other reportable incidents	47265	Dual-chamber implantable pacemaker, rate-responsive	CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
12581	2023	All other reportable incidents	17882	Physiologic-monitoring defibrillation system	CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A0719	Unexpected Shutdown		
							F27	Problem Identified During Non-Clinical Procedure	A0719	Unexpected Shutdown		
12582	2023	All other reportable incidents	47265	Dual-chamber implantable pacemaker, rate-responsive	CT141	Electrodes/Leads and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	10/03/2023	10/03/2023
							F1901	Additional Surgery	A070504	Premature Discharge of Battery	10/03/2023	10/03/2023
					CT145	Cardiovascular devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	10/03/2023	10/03/2023
							F1901	Additional Surgery	A070504	Premature Discharge of Battery	10/03/2023	10/03/2023
					CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	10/03/2023	10/03/2023
							F1901	Additional Surgery	A070504	Premature Discharge of Battery	10/03/2023	10/03/2023
					CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	10/03/2023	10/03/2023
							F1901	Additional Surgery	A070504	Premature Discharge of Battery	10/03/2023	10/03/2023
					CT233	Surgical	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	10/03/2023	10/03/2023
							F1901	Additional Surgery	A070504	Premature Discharge of Battery	10/03/2023	10/03/2023
					CT2406	Implantable	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	10/03/2023	10/03/2023
							F1901	Additional Surgery	A070504	Premature Discharge of Battery	10/03/2023	10/03/2023
					CT2407	Active implantable	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	10/03/2023	10/03/2023
							F1901	Additional Surgery	A070504	Premature Discharge of Battery	10/03/2023	10/03/2023
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	10/03/2023	10/03/2023
							F1901	Additional Surgery	A070504	Premature Discharge of Battery	10/03/2023	10/03/2023
					CT2727	Active implant devices/ancillaries	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	10/03/2023	10/03/2023
							F1901	Additional Surgery	A070504	Premature Discharge of Battery	10/03/2023	10/03/2023
					CT2750	Implantable cardiac electrical stimulators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	10/03/2023	10/03/2023
							F1901	Additional Surgery	A070504	Premature Discharge of Battery	10/03/2023	10/03/2023
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	10/03/2023	10/03/2023
							F1901	Additional Surgery	A070504	Premature Discharge of Battery	10/03/2023	10/03/2023
					CT2985	Active	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	10/03/2023	10/03/2023
							F1901	Additional Surgery	A070504	Premature Discharge of Battery	10/03/2023	10/03/2023
					CT3001	Implantable metal	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	10/03/2023	10/03/2023
							F1901	Additional Surgery	A070504	Premature Discharge of Battery	10/03/2023	10/03/2023
					CT3002	Implantable synthetic polymer	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	10/03/2023	10/03/2023
							F1901	Additional Surgery	A070504	Premature Discharge of Battery	10/03/2023	10/03/2023
					CT3006	Implantable inorganic material	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	10/03/2023	10/03/2023
							F1901	Additional Surgery	A070504	Premature Discharge of Battery	10/03/2023	10/03/2023
					CT314	Electrically powered	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	10/03/2023	10/03/2023
							F1901	Additional Surgery	A070504	Premature Discharge of Battery	10/03/2023	10/03/2023
					CT321	Long-term surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	10/03/2023	10/03/2023
							F1901	Additional Surgery	A070504	Premature Discharge of Battery	10/03/2023	10/03/2023
					CT449	Generators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	10/03/2023	10/03/2023
							F1901	Additional Surgery	A070504	Premature Discharge of Battery	10/03/2023	10/03/2023
					CT981	Single use	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	10/03/2023	10/03/2023
							F1901	Additional Surgery	A070504	Premature Discharge of Battery	10/03/2023	10/03/2023
					CT983	Surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	10/03/2023	10/03/2023
							F1901	Additional Surgery	A070504	Premature Discharge of Battery	10/03/2023	10/03/2023
12583	2023	All other reportable incidents	47265	Dual-chamber implantable pacemaker, rate-responsive	CT141	Electrodes/Leads and associated devices	E0112	Dizziness	A070504	Premature Discharge of Battery	02/03/2023	02/03/2023
									A071204	No Pacing	02/03/2023	02/03/2023
							F1901	Additional Surgery	A070504	Premature Discharge of Battery	02/03/2023	02/03/2023
									A071204	No Pacing	02/03/2023	02/03/2023
					CT145	Cardiovascular devices	E0112	Dizziness	A070504	Premature Discharge of Battery	02/03/2023	02/03/2023
									A071204	No Pacing	02/03/2023	02/03/2023
							F1901	Additional Surgery	A070504	Premature Discharge of Battery	02/03/2023	02/03/2023
									A071204	No Pacing	02/03/2023	02/03/2023
					CT1684	Cardiac pacemakers and associated devices	E0112	Dizziness	A070504	Premature Discharge of Battery	02/03/2023	02/03/2023
									A071204	No Pacing	02/03/2023	02/03/2023
							F1901	Additional Surgery	A070504	Premature Discharge of Battery	02/03/2023	02/03/2023
									A071204	No Pacing	02/03/2023	02/03/2023
					CT1685	Cardiac pacemakers	E0112	Dizziness	A070504	Premature Discharge of Battery	02/03/2023	02/03/2023
									A071204	No Pacing	02/03/2023	02/03/2023
							F1901	Additional Surgery	A070504	Premature Discharge of Battery	02/03/2023	02/03/2023
									A071204	No Pacing	02/03/2023	02/03/2023
					CT233	Surgical	E0112	Dizziness	A070504	Premature Discharge of Battery	02/03/2023	02/03/2023
									A071204	No Pacing	02/03/2023	02/03/2023
							F1901	Additional Surgery	A070504	Premature Discharge of Battery	02/03/2023	02/03/2023
									A071204	No Pacing	02/03/2023	02/03/2023
					CT2406	Implantable	E0112	Dizziness	A070504	Premature Discharge of Battery	02/03/2023	02/03/2023
									A071204	No Pacing	02/03/2023	02/03/2023
							F1901	Additional Surgery	A070504	Premature Discharge of Battery	02/03/2023	02/03/2023
									A071204	No Pacing	02/03/2023	02/03/2023
					CT2407	Active implantable	E0112	Dizziness	A070504	Premature Discharge of Battery	02/03/2023	02/03/2023
									A071204	No Pacing	02/03/2023	02/03/2023
							F1901	Additional Surgery	A070504	Premature Discharge of Battery	02/03/2023	02/03/2023
									A071204	No Pacing	02/03/2023	02/03/2023
					CT257	Implantable cardiac pacemakers	E0112	Dizziness	A070504	Premature Discharge of Battery	02/03/2023	02/03/2023
									A071204	No Pacing	02/03/2023	02/03/2023
							F1901	Additional Surgery	A070504	Premature Discharge of Battery	02/03/2023	02/03/2023
									A071204	No Pacing	02/03/2023	02/03/2023
					CT2727	Active implant devices/ancillaries	E0112	Dizziness	A070504	Premature Discharge of Battery	02/03/2023	02/03/2023
									A071204	No Pacing	02/03/2023	02/03/2023
							F1901	Additional Surgery	A070504	Premature Discharge of Battery	02/03/2023	02/03/2023
									A071204	No Pacing	02/03/2023	02/03/2023
					CT2750	Implantable cardiac electrical stimulators and associated devices	E0112	Dizziness	A070504	Premature Discharge of Battery	02/03/2023	02/03/2023
									A071204	No Pacing	02/03/2023	02/03/2023
							F1901	Additional Surgery	A070504	Premature Discharge of Battery	02/03/2023	02/03/2023
									A071204	No Pacing	02/03/2023	02/03/2023

					CT2866	Implantable cardiac electrical stimulators	E0112	Dizziness	A070504	Premature Discharge of Battery	02/03/2023	02/03/2023
									A071204	No Pacing	02/03/2023	02/03/2023
							F1901	Additional Surgery	A070504	Premature Discharge of Battery	02/03/2023	02/03/2023
									A071204	No Pacing	02/03/2023	02/03/2023
					CT2985	Active	E0112	Dizziness	A070504	Premature Discharge of Battery	02/03/2023	02/03/2023
									A071204	No Pacing	02/03/2023	02/03/2023
							F1901	Additional Surgery	A070504	Premature Discharge of Battery	02/03/2023	02/03/2023
									A071204	No Pacing	02/03/2023	02/03/2023
					CT3001	Implantable metal	E0112	Dizziness	A070504	Premature Discharge of Battery	02/03/2023	02/03/2023
									A071204	No Pacing	02/03/2023	02/03/2023
							F1901	Additional Surgery	A070504	Premature Discharge of Battery	02/03/2023	02/03/2023
									A071204	No Pacing	02/03/2023	02/03/2023
					CT3002	Implantable synthetic polymer	E0112	Dizziness	A070504	Premature Discharge of Battery	02/03/2023	02/03/2023
									A071204	No Pacing	02/03/2023	02/03/2023
							F1901	Additional Surgery	A070504	Premature Discharge of Battery	02/03/2023	02/03/2023
									A071204	No Pacing	02/03/2023	02/03/2023
					CT3006	Implantable inorganic material	E0112	Dizziness	A070504	Premature Discharge of Battery	02/03/2023	02/03/2023
									A071204	No Pacing	02/03/2023	02/03/2023
							F1901	Additional Surgery	A070504	Premature Discharge of Battery	02/03/2023	02/03/2023
									A071204	No Pacing	02/03/2023	02/03/2023
					CT314	Electrically powered	E0112	Dizziness	A070504	Premature Discharge of Battery	02/03/2023	02/03/2023
									A071204	No Pacing	02/03/2023	02/03/2023
							F1901	Additional Surgery	A070504	Premature Discharge of Battery	02/03/2023	02/03/2023
									A071204	No Pacing	02/03/2023	02/03/2023
					CT321	Long-term surgical invasive	E0112	Dizziness	A070504	Premature Discharge of Battery	02/03/2023	02/03/2023
									A071204	No Pacing	02/03/2023	02/03/2023
							F1901	Additional Surgery	A070504	Premature Discharge of Battery	02/03/2023	02/03/2023
									A071204	No Pacing	02/03/2023	02/03/2023
					CT449	Generators and associated devices	E0112	Dizziness	A070504	Premature Discharge of Battery	02/03/2023	02/03/2023
									A071204	No Pacing	02/03/2023	02/03/2023
							F1901	Additional Surgery	A070504	Premature Discharge of Battery	02/03/2023	02/03/2023
									A071204	No Pacing	02/03/2023	02/03/2023
					CT981	Single use	E0112	Dizziness	A070504	Premature Discharge of Battery	02/03/2023	02/03/2023
									A071204	No Pacing	02/03/2023	02/03/2023
							F1901	Additional Surgery	A070504	Premature Discharge of Battery	02/03/2023	02/03/2023
									A071204	No Pacing	02/03/2023	02/03/2023
					CT983	Surgical invasive	E0112	Dizziness	A070504	Premature Discharge of Battery	02/03/2023	02/03/2023
									A071204	No Pacing	02/03/2023	02/03/2023
							F1901	Additional Surgery	A070504	Premature Discharge of Battery	02/03/2023	02/03/2023
									A071204	No Pacing	02/03/2023	02/03/2023
12584	2023	All other reportable incidents	47265	Dual-chamber implantable pacemaker, rate-responsive	CT141	Electrodes/Leads and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	08/02/2023	08/02/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	08/02/2023	08/02/2023
					CT145	Cardiovascular devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	08/02/2023	08/02/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	08/02/2023	08/02/2023
					CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	08/02/2023	08/02/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	08/02/2023	08/02/2023
					CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	08/02/2023	08/02/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	08/02/2023	08/02/2023
					CT233	Surgical	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	08/02/2023	08/02/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	08/02/2023	08/02/2023
					CT2406	Implantable	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	08/02/2023	08/02/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	08/02/2023	08/02/2023
					CT2407	Active implantable	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	08/02/2023	08/02/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	08/02/2023	08/02/2023
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	08/02/2023	08/02/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	08/02/2023	08/02/2023
					CT2727	Active implant devices/ancillaries	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	08/02/2023	08/02/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	08/02/2023	08/02/2023
					CT2750	Implantable cardiac electrical stimulators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	08/02/2023	08/02/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	08/02/2023	08/02/2023
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	08/02/2023	08/02/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	08/02/2023	08/02/2023
					CT2985	Active	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	08/02/2023	08/02/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	08/02/2023	08/02/2023
					CT3001	Implantable metal	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	08/02/2023	08/02/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	08/02/2023	08/02/2023
					CT3002	Implantable synthetic polymer	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	08/02/2023	08/02/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	08/02/2023	08/02/2023
					CT3006	Implantable inorganic material	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	08/02/2023	08/02/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	08/02/2023	08/02/2023
					CT314	Electrically powered	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	08/02/2023	08/02/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	08/02/2023	08/02/2023
					CT321	Long-term surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	08/02/2023	08/02/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	08/02/2023	08/02/2023
					CT449	Generators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	08/02/2023	08/02/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	08/02/2023	08/02/2023
					CT981	Single use	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	08/02/2023	08/02/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	08/02/2023	08/02/2023
					CT983	Surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	08/02/2023	08/02/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	08/02/2023	08/02/2023
12585	2023		47265	Dual-chamber implantable pacemaker, rate-responsive	CT141	Electrodes/Leads and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	10/03/2023	10/03/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	10/03/2023	10/03/2023
					CT145	Cardiovascular devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	10/03/2023	10/03/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	10/03/2023	10/03/2023
					CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	10/03/2023	10/03/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	10/03/2023	10/03/2023
					CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	10/03/2023	10/03/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	10/03/2023	10/03/2023

					CT233	Surgical	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	10/03/2023	10/03/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	10/03/2023	10/03/2023
					CT2406	Implantable	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	10/03/2023	10/03/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	10/03/2023	10/03/2023
					CT2407	Active implantable	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	10/03/2023	10/03/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	10/03/2023	10/03/2023
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	10/03/2023	10/03/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	10/03/2023	10/03/2023
					CT2727	Active implant devices/ancillaries	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	10/03/2023	10/03/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	10/03/2023	10/03/2023
					CT2750	Implantable cardiac electrical stimulators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	10/03/2023	10/03/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	10/03/2023	10/03/2023
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	10/03/2023	10/03/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	10/03/2023	10/03/2023
					CT2985	Active	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	10/03/2023	10/03/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	10/03/2023	10/03/2023
					CT3001	Implantable metal	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	10/03/2023	10/03/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	10/03/2023	10/03/2023
					CT3002	Implantable synthetic polymer	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	10/03/2023	10/03/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	10/03/2023	10/03/2023
					CT3006	Implantable inorganic material	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	10/03/2023	10/03/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	10/03/2023	10/03/2023
					CT314	Electrically powered	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	10/03/2023	10/03/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	10/03/2023	10/03/2023
					CT321	Long-term surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	10/03/2023	10/03/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	10/03/2023	10/03/2023
					CT449	Generators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	10/03/2023	10/03/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	10/03/2023	10/03/2023
					CT981	Single use	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	10/03/2023	10/03/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	10/03/2023	10/03/2023
					CT983	Surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	10/03/2023	10/03/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	10/03/2023	10/03/2023
12586	2023	All other reportable incidents	35852	Single-chamber implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
									A0908	Incorrect, Inadequate or Imprecise Result or Readings		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
									A0908	Incorrect, Inadequate or Imprecise Result or Readings		
12587	2023	All other reportable incidents	35852	Single-chamber implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
12588	2023	All other reportable incidents	47265	Dual-chamber implantable pacemaker, rate-responsive	CT141	Electrodes/Leads and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT145	Cardiovascular devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT233	Surgical	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT2406	Implantable	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT2407	Active implantable	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT2727	Active implant devices/ancillaries	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT2750	Implantable cardiac electrical stimulators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT2985	Active	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT3001	Implantable metal	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT3002	Implantable synthetic polymer	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT3006	Implantable inorganic material	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT314	Electrically powered	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT321	Long-term surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT449	Generators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT981	Single use	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT983	Surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
12589	2023	All other reportable incidents	47265	Dual-chamber implantable pacemaker, rate-responsive	CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	25/05/2023	25/05/2023
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	25/05/2023	25/05/2023
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	25/05/2023	25/05/2023
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	25/05/2023	25/05/2023
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	25/05/2023	25/05/2023
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	25/05/2023	25/05/2023

					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	25/05/2023	25/05/2023
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	25/05/2023	25/05/2023
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	25/05/2023	25/05/2023
12590	2023	All other reportable incidents	47265	Dual-chamber implantable pacemaker, rate-responsive	CT141	Electrodes/Leads and associated devices	E0601	Arrhythmia	A0712	Pacing Problem		
							F08	Hospitalization or Prolonged Hospitalization	A0712	Pacing Problem		
					CT145	Cardiovascular devices	E0601	Arrhythmia	A0712	Pacing Problem		
							F08	Hospitalization or Prolonged Hospitalization	A0712	Pacing Problem		
					CT1684	Cardiac pacemakers and associated devices	E0601	Arrhythmia	A0712	Pacing Problem		
							F08	Hospitalization or Prolonged Hospitalization	A0712	Pacing Problem		
					CT1685	Cardiac pacemakers	E0601	Arrhythmia	A0712	Pacing Problem		
							F08	Hospitalization or Prolonged Hospitalization	A0712	Pacing Problem		
					CT233	Surgical	E0601	Arrhythmia	A0712	Pacing Problem		
							F08	Hospitalization or Prolonged Hospitalization	A0712	Pacing Problem		
					CT2406	Implantable	E0601	Arrhythmia	A0712	Pacing Problem		
							F08	Hospitalization or Prolonged Hospitalization	A0712	Pacing Problem		
					CT2407	Active implantable	E0601	Arrhythmia	A0712	Pacing Problem		
							F08	Hospitalization or Prolonged Hospitalization	A0712	Pacing Problem		
					CT257	Implantable cardiac pacemakers	E0601	Arrhythmia	A0712	Pacing Problem		
							F08	Hospitalization or Prolonged Hospitalization	A0712	Pacing Problem		
					CT2727	Active implant devices/ancillaries	E0601	Arrhythmia	A0712	Pacing Problem		
							F08	Hospitalization or Prolonged Hospitalization	A0712	Pacing Problem		
					CT2750	Implantable cardiac electrical stimulators and associated devices	E0601	Arrhythmia	A0712	Pacing Problem		
							F08	Hospitalization or Prolonged Hospitalization	A0712	Pacing Problem		
					CT2866	Implantable cardiac electrical stimulators	E0601	Arrhythmia	A0712	Pacing Problem		
							F08	Hospitalization or Prolonged Hospitalization	A0712	Pacing Problem		
					CT2985	Active	E0601	Arrhythmia	A0712	Pacing Problem		
							F08	Hospitalization or Prolonged Hospitalization	A0712	Pacing Problem		
					CT3001	Implantable metal	E0601	Arrhythmia	A0712	Pacing Problem		
							F08	Hospitalization or Prolonged Hospitalization	A0712	Pacing Problem		
					CT3002	Implantable synthetic polymer	E0601	Arrhythmia	A0712	Pacing Problem		
							F08	Hospitalization or Prolonged Hospitalization	A0712	Pacing Problem		
					CT3006	Implantable inorganic material	E0601	Arrhythmia	A0712	Pacing Problem		
							F08	Hospitalization or Prolonged Hospitalization	A0712	Pacing Problem		
					CT314	Electrically powered	E0601	Arrhythmia	A0712	Pacing Problem		
							F08	Hospitalization or Prolonged Hospitalization	A0712	Pacing Problem		
					CT321	Long-term surgical invasive	E0601	Arrhythmia	A0712	Pacing Problem		
							F08	Hospitalization or Prolonged Hospitalization	A0712	Pacing Problem		
					CT449	Generators and associated devices	E0601	Arrhythmia	A0712	Pacing Problem		
							F08	Hospitalization or Prolonged Hospitalization	A0712	Pacing Problem		
					CT981	Single use	E0601	Arrhythmia	A0712	Pacing Problem		
							F08	Hospitalization or Prolonged Hospitalization	A0712	Pacing Problem		
					CT983	Surgical invasive	E0601	Arrhythmia	A0712	Pacing Problem		
							F08	Hospitalization or Prolonged Hospitalization	A0712	Pacing Problem		
12591	2023	All other reportable incidents	47265	Dual-chamber implantable pacemaker, rate-responsive	CT141	Electrodes/Leads and associated devices	E060106	Heart Block	A070101	Failure to Capture	23/02/2023	23/02/2023
									A070102	High Capture Threshold	23/02/2023	23/02/2023
									A070901	Decreased Sensitivity	23/02/2023	23/02/2023
									A072201	High Impedance	23/02/2023	23/02/2023
							F08	Hospitalization or Prolonged Hospitalization	A070101	Failure to Capture	23/02/2023	23/02/2023
									A070102	High Capture Threshold	23/02/2023	23/02/2023
									A070901	Decreased Sensitivity	23/02/2023	23/02/2023
									A072201	High Impedance	23/02/2023	23/02/2023
							F1905	Device Revision or Replacement	A070101	Failure to Capture	23/02/2023	23/02/2023
									A070102	High Capture Threshold	23/02/2023	23/02/2023
									A070901	Decreased Sensitivity	23/02/2023	23/02/2023
									A072201	High Impedance	23/02/2023	23/02/2023
					CT145	Cardiovascular devices	E060106	Heart Block	A070101	Failure to Capture	23/02/2023	23/02/2023
									A070102	High Capture Threshold	23/02/2023	23/02/2023
									A070901	Decreased Sensitivity	23/02/2023	23/02/2023
									A072201	High Impedance	23/02/2023	23/02/2023
							F08	Hospitalization or Prolonged Hospitalization	A070101	Failure to Capture	23/02/2023	23/02/2023
									A070102	High Capture Threshold	23/02/2023	23/02/2023
									A070901	Decreased Sensitivity	23/02/2023	23/02/2023
									A072201	High Impedance	23/02/2023	23/02/2023
							F1905	Device Revision or Replacement	A070101	Failure to Capture	23/02/2023	23/02/2023
									A070102	High Capture Threshold	23/02/2023	23/02/2023
									A070901	Decreased Sensitivity	23/02/2023	23/02/2023
									A072201	High Impedance	23/02/2023	23/02/2023
					CT1684	Cardiac pacemakers and associated devices	E060106	Heart Block	A070101	Failure to Capture	23/02/2023	23/02/2023
									A070102	High Capture Threshold	23/02/2023	23/02/2023
									A070901	Decreased Sensitivity	23/02/2023	23/02/2023
									A072201	High Impedance	23/02/2023	23/02/2023
							F08	Hospitalization or Prolonged Hospitalization	A070101	Failure to Capture	23/02/2023	23/02/2023
									A070102	High Capture Threshold	23/02/2023	23/02/2023
									A070901	Decreased Sensitivity	23/02/2023	23/02/2023
									A072201	High Impedance	23/02/2023	23/02/2023
							F1905	Device Revision or Replacement	A070101	Failure to Capture	23/02/2023	23/02/2023
									A070102	High Capture Threshold	23/02/2023	23/02/2023
									A070901	Decreased Sensitivity	23/02/2023	23/02/2023
									A072201	High Impedance	23/02/2023	23/02/2023
					CT1685	Cardiac pacemakers	E060106	Heart Block	A070101	Failure to Capture	23/02/2023	23/02/2023
									A070102	High Capture Threshold	23/02/2023	23/02/2023
									A070901	Decreased Sensitivity	23/02/2023	23/02/2023
									A072201	High Impedance	23/02/2023	23/02/2023
							F08	Hospitalization or Prolonged Hospitalization	A070101	Failure to Capture	23/02/2023	23/02/2023
									A070102	High Capture Threshold	23/02/2023	23/02/2023
									A070901	Decreased Sensitivity	23/02/2023	23/02/2023
									A072201	High Impedance	23/02/2023	23/02/2023
							F1905	Device Revision or Replacement	A070101	Failure to Capture	23/02/2023	23/02/2023

								A070102	High Capture Threshold	23/02/2023	23/02/2023
								A070901	Decreased Sensitivity	23/02/2023	23/02/2023
								A072201	High Impedance	23/02/2023	23/02/2023
					CT233	Surgical	E060106	Heart Block	A070101	Failure to Capture	23/02/2023	23/02/2023
									A070102	High Capture Threshold	23/02/2023	23/02/2023
									A070901	Decreased Sensitivity	23/02/2023	23/02/2023
									A072201	High Impedance	23/02/2023	23/02/2023
							F08	Hospitalization or Prolonged Hospitalization	A070101	Failure to Capture	23/02/2023	23/02/2023
									A070102	High Capture Threshold	23/02/2023	23/02/2023
									A070901	Decreased Sensitivity	23/02/2023	23/02/2023
									A072201	High Impedance	23/02/2023	23/02/2023
							F1905	Device Revision or Replacement	A070101	Failure to Capture	23/02/2023	23/02/2023
									A070102	High Capture Threshold	23/02/2023	23/02/2023
									A070901	Decreased Sensitivity	23/02/2023	23/02/2023
									A072201	High Impedance	23/02/2023	23/02/2023
					CT2406	Implantable	E060106	Heart Block	A070101	Failure to Capture	23/02/2023	23/02/2023
									A070102	High Capture Threshold	23/02/2023	23/02/2023
									A070901	Decreased Sensitivity	23/02/2023	23/02/2023
									A072201	High Impedance	23/02/2023	23/02/2023
							F08	Hospitalization or Prolonged Hospitalization	A070101	Failure to Capture	23/02/2023	23/02/2023
									A070102	High Capture Threshold	23/02/2023	23/02/2023
									A070901	Decreased Sensitivity	23/02/2023	23/02/2023
									A072201	High Impedance	23/02/2023	23/02/2023
							F1905	Device Revision or Replacement	A070101	Failure to Capture	23/02/2023	23/02/2023
									A070102	High Capture Threshold	23/02/2023	23/02/2023
									A070901	Decreased Sensitivity	23/02/2023	23/02/2023
									A072201	High Impedance	23/02/2023	23/02/2023
					CT2407	Active implantable	E060106	Heart Block	A070101	Failure to Capture	23/02/2023	23/02/2023
									A070102	High Capture Threshold	23/02/2023	23/02/2023
									A070901	Decreased Sensitivity	23/02/2023	23/02/2023
									A072201	High Impedance	23/02/2023	23/02/2023
							F08	Hospitalization or Prolonged Hospitalization	A070101	Failure to Capture	23/02/2023	23/02/2023
									A070102	High Capture Threshold	23/02/2023	23/02/2023
									A070901	Decreased Sensitivity	23/02/2023	23/02/2023
									A072201	High Impedance	23/02/2023	23/02/2023
							F1905	Device Revision or Replacement	A070101	Failure to Capture	23/02/2023	23/02/2023
									A070102	High Capture Threshold	23/02/2023	23/02/2023
									A070901	Decreased Sensitivity	23/02/2023	23/02/2023
									A072201	High Impedance	23/02/2023	23/02/2023
					CT257	Implantable cardiac pacemakers	E060106	Heart Block	A070101	Failure to Capture	23/02/2023	23/02/2023
									A070102	High Capture Threshold	23/02/2023	23/02/2023
									A070901	Decreased Sensitivity	23/02/2023	23/02/2023
									A072201	High Impedance	23/02/2023	23/02/2023
							F08	Hospitalization or Prolonged Hospitalization	A070101	Failure to Capture	23/02/2023	23/02/2023
									A070102	High Capture Threshold	23/02/2023	23/02/2023
									A070901	Decreased Sensitivity	23/02/2023	23/02/2023
									A072201	High Impedance	23/02/2023	23/02/2023
							F1905	Device Revision or Replacement	A070101	Failure to Capture	23/02/2023	23/02/2023
									A070102	High Capture Threshold	23/02/2023	23/02/2023
									A070901	Decreased Sensitivity	23/02/2023	23/02/2023
									A072201	High Impedance	23/02/2023	23/02/2023
					CT2727	Active implant devices/ancillaries	E060106	Heart Block	A070101	Failure to Capture	23/02/2023	23/02/2023
									A070102	High Capture Threshold	23/02/2023	23/02/2023
									A070901	Decreased Sensitivity	23/02/2023	23/02/2023
									A072201	High Impedance	23/02/2023	23/02/2023
							F08	Hospitalization or Prolonged Hospitalization	A070101	Failure to Capture	23/02/2023	23/02/2023
									A070102	High Capture Threshold	23/02/2023	23/02/2023
									A070901	Decreased Sensitivity	23/02/2023	23/02/2023
									A072201	High Impedance	23/02/2023	23/02/2023
							F1905	Device Revision or Replacement	A070101	Failure to Capture	23/02/2023	23/02/2023
									A070102	High Capture Threshold	23/02/2023	23/02/2023
									A070901	Decreased Sensitivity	23/02/2023	23/02/2023
									A072201	High Impedance	23/02/2023	23/02/2023
					CT2750	Implantable cardiac electrical stimulators and associated devices	E060106	Heart Block	A070101	Failure to Capture	23/02/2023	23/02/2023
									A070102	High Capture Threshold	23/02/2023	23/02/2023
									A070901	Decreased Sensitivity	23/02/2023	23/02/2023
									A072201	High Impedance	23/02/2023	23/02/2023
							F08	Hospitalization or Prolonged Hospitalization	A070101	Failure to Capture	23/02/2023	23/02/2023
									A070102	High Capture Threshold	23/02/2023	23/02/2023
									A070901	Decreased Sensitivity	23/02/2023	23/02/2023
									A072201	High Impedance	23/02/2023	23/02/2023
							F1905	Device Revision or Replacement	A070101	Failure to Capture	23/02/2023	23/02/2023
									A070102	High Capture Threshold	23/02/2023	23/02/2023
									A070901	Decreased Sensitivity	23/02/2023	23/02/2023
									A072201	High Impedance	23/02/2023	23/02/2023
					CT2866	Implantable cardiac electrical stimulators	E060106	Heart Block	A070101	Failure to Capture	23/02/2023	23/02/2023
									A070102	High Capture Threshold	23/02/2023	23/02/2023
									A070901	Decreased Sensitivity	23/02/2023	23/02/2023
									A072201	High Impedance	23/02/2023	23/02/2023
							F08	Hospitalization or Prolonged Hospitalization	A070101	Failure to Capture	23/02/2023	23/02/2023
									A070102	High Capture Threshold	23/02/2023	23/02/2023
									A070901	Decreased Sensitivity	23/02/2023	23/02/2023
									A072201	High Impedance	23/02/2023	23/02/2023
							F1905	Device Revision or Replacement	A070101	Failure to Capture	23/02/2023	23/02/2023
									A070102	High Capture Threshold	23/02/2023	23/02/2023
									A070901	Decreased Sensitivity	23/02/2023	23/02/2023
									A072201	High Impedance	23/02/2023	23/02/2023
					CT2985	Active	E060106	Heart Block	A070101	Failure to Capture	23/02/2023	23/02/2023

									A070102	High Capture Threshold	23/02/2023	23/02/2023
									A070901	Decreased Sensitivity	23/02/2023	23/02/2023
									A072201	High Impedance	23/02/2023	23/02/2023
								F08	Hospitalization or Prolonged Hospitalization	A070101	Failure to Capture	23/02/2023	23/02/2023
										A070102	High Capture Threshold	23/02/2023	23/02/2023
										A070901	Decreased Sensitivity	23/02/2023	23/02/2023
										A072201	High Impedance	23/02/2023	23/02/2023
								F1905	Device Revision or Replacement	A070101	Failure to Capture	23/02/2023	23/02/2023
										A070102	High Capture Threshold	23/02/2023	23/02/2023
										A070901	Decreased Sensitivity	23/02/2023	23/02/2023
										A072201	High Impedance	23/02/2023	23/02/2023
						CT3001	Implantable metal	E060106	Heart Block	A070101	Failure to Capture	23/02/2023	23/02/2023
										A070102	High Capture Threshold	23/02/2023	23/02/2023
										A070901	Decreased Sensitivity	23/02/2023	23/02/2023
										A072201	High Impedance	23/02/2023	23/02/2023
								F08	Hospitalization or Prolonged Hospitalization	A070101	Failure to Capture	23/02/2023	23/02/2023
										A070102	High Capture Threshold	23/02/2023	23/02/2023
										A070901	Decreased Sensitivity	23/02/2023	23/02/2023
										A072201	High Impedance	23/02/2023	23/02/2023
								F1905	Device Revision or Replacement	A070101	Failure to Capture	23/02/2023	23/02/2023
										A070102	High Capture Threshold	23/02/2023	23/02/2023
										A070901	Decreased Sensitivity	23/02/2023	23/02/2023
										A072201	High Impedance	23/02/2023	23/02/2023
						CT3002	Implantable synthetic polymer	E060106	Heart Block	A070101	Failure to Capture	23/02/2023	23/02/2023
										A070102	High Capture Threshold	23/02/2023	23/02/2023
										A070901	Decreased Sensitivity	23/02/2023	23/02/2023
										A072201	High Impedance	23/02/2023	23/02/2023
								F08	Hospitalization or Prolonged Hospitalization	A070101	Failure to Capture	23/02/2023	23/02/2023
										A070102	High Capture Threshold	23/02/2023	23/02/2023
										A070901	Decreased Sensitivity	23/02/2023	23/02/2023
										A072201	High Impedance	23/02/2023	23/02/2023
								F1905	Device Revision or Replacement	A070101	Failure to Capture	23/02/2023	23/02/2023
										A070102	High Capture Threshold	23/02/2023	23/02/2023
										A070901	Decreased Sensitivity	23/02/2023	23/02/2023
										A072201	High Impedance	23/02/2023	23/02/2023
						CT3006	Implantable inorganic material	E060106	Heart Block	A070101	Failure to Capture	23/02/2023	23/02/2023
										A070102	High Capture Threshold	23/02/2023	23/02/2023
										A070901	Decreased Sensitivity	23/02/2023	23/02/2023
										A072201	High Impedance	23/02/2023	23/02/2023
								F08	Hospitalization or Prolonged Hospitalization	A070101	Failure to Capture	23/02/2023	23/02/2023
										A070102	High Capture Threshold	23/02/2023	23/02/2023
										A070901	Decreased Sensitivity	23/02/2023	23/02/2023
										A072201	High Impedance	23/02/2023	23/02/2023
								F1905	Device Revision or Replacement	A070101	Failure to Capture	23/02/2023	23/02/2023
										A070102	High Capture Threshold	23/02/2023	23/02/2023
										A070901	Decreased Sensitivity	23/02/2023	23/02/2023
										A072201	High Impedance	23/02/2023	23/02/2023
						CT314	Electrically powered	E060106	Heart Block	A070101	Failure to Capture	23/02/2023	23/02/2023
										A070102	High Capture Threshold	23/02/2023	23/02/2023
										A070901	Decreased Sensitivity	23/02/2023	23/02/2023
										A072201	High Impedance	23/02/2023	23/02/2023
								F08	Hospitalization or Prolonged Hospitalization	A070101	Failure to Capture	23/02/2023	23/02/2023
										A070102	High Capture Threshold	23/02/2023	23/02/2023
										A070901	Decreased Sensitivity	23/02/2023	23/02/2023
										A072201	High Impedance	23/02/2023	23/02/2023
								F1905	Device Revision or Replacement	A070101	Failure to Capture	23/02/2023	23/02/2023
										A070102	High Capture Threshold	23/02/2023	23/02/2023
										A070901	Decreased Sensitivity	23/02/2023	23/02/2023
										A072201	High Impedance	23/02/2023	23/02/2023
						CT321	Long-term surgical invasive	E060106	Heart Block	A070101	Failure to Capture	23/02/2023	23/02/2023
										A070102	High Capture Threshold	23/02/2023	23/02/2023
										A070901	Decreased Sensitivity	23/02/2023	23/02/2023
										A072201	High Impedance	23/02/2023	23/02/2023
								F08	Hospitalization or Prolonged Hospitalization	A070101	Failure to Capture	23/02/2023	23/02/2023
										A070102	High Capture Threshold	23/02/2023	23/02/2023
										A070901	Decreased Sensitivity	23/02/2023	23/02/2023
										A072201	High Impedance	23/02/2023	23/02/2023
								F1905	Device Revision or Replacement	A070101	Failure to Capture	23/02/2023	23/02/2023
										A070102	High Capture Threshold	23/02/2023	23/02/2023
										A070901	Decreased Sensitivity	23/02/2023	23/02/2023
										A072201	High Impedance	23/02/2023	23/02/2023
						CT449	Generators and associated devices	E060106	Heart Block	A070101	Failure to Capture	23/02/2023	23/02/2023
										A070102	High Capture Threshold	23/02/2023	23/02/2023
										A070901	Decreased Sensitivity	23/02/2023	23/02/2023
										A072201	High Impedance	23/02/2023	23/02/2023
								F08	Hospitalization or Prolonged Hospitalization	A070101	Failure to Capture	23/02/2023	23/02/2023
										A070102	High Capture Threshold	23/02/2023	23/02/2023
										A070901	Decreased Sensitivity	23/02/2023	23/02/2023
										A072201	High Impedance	23/02/2023	23/02/2023
								F1905	Device Revision or Replacement	A070101	Failure to Capture	23/02/2023	23/02/2023
										A070102	High Capture Threshold	23/02/2023	23/02/2023
										A070901	Decreased Sensitivity	23/02/2023	23/02/2023
										A072201	High Impedance	23/02/2023	23/02/2023
						CT981	Single use	E060106	Heart Block	A070101	Failure to Capture	23/02/2023	23/02/2023
										A070102	High Capture Threshold	23/02/2023	23/02/2023
										A070901	Decreased Sensitivity	23/02/2023	23/02/2023
										A072201	High Impedance	23/02/2023	23/02/2023
								F08	Hospitalization or Prolonged Hospitalization	A070101	Failure to Capture	23/02/2023	23/02/2023

									A070102	High Capture Threshold	23/02/2023	23/02/2023
									A070901	Decreased Sensitivity	23/02/2023	23/02/2023
									A072201	High Impedance	23/02/2023	23/02/2023
							F1905	Device Revision or Replacement	A070101	Failure to Capture	23/02/2023	23/02/2023
									A070102	High Capture Threshold	23/02/2023	23/02/2023
									A070901	Decreased Sensitivity	23/02/2023	23/02/2023
									A072201	High Impedance	23/02/2023	23/02/2023
					CT983	Surgical invasive	E060106	Heart Block	A070101	Failure to Capture	23/02/2023	23/02/2023
									A070102	High Capture Threshold	23/02/2023	23/02/2023
									A070901	Decreased Sensitivity	23/02/2023	23/02/2023
									A072201	High Impedance	23/02/2023	23/02/2023
							F08	Hospitalization or Prolonged Hospitalization	A070101	Failure to Capture	23/02/2023	23/02/2023
									A070102	High Capture Threshold	23/02/2023	23/02/2023
									A070901	Decreased Sensitivity	23/02/2023	23/02/2023
									A072201	High Impedance	23/02/2023	23/02/2023
							F1905	Device Revision or Replacement	A070101	Failure to Capture	23/02/2023	23/02/2023
									A070102	High Capture Threshold	23/02/2023	23/02/2023
									A070901	Decreased Sensitivity	23/02/2023	23/02/2023
									A072201	High Impedance	23/02/2023	23/02/2023
12592	2023	All other reportable incidents	47265	Dual-chamber implantable pacemaker, rate-responsive	CT141	Electrodes/Leads and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	08/03/2023	08/03/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	08/03/2023	08/03/2023
					CT145	Cardiovascular devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	08/03/2023	08/03/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	08/03/2023	08/03/2023
					CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	08/03/2023	08/03/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	08/03/2023	08/03/2023
					CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	08/03/2023	08/03/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	08/03/2023	08/03/2023
					CT233	Surgical	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	08/03/2023	08/03/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	08/03/2023	08/03/2023
					CT2406	Implantable	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	08/03/2023	08/03/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	08/03/2023	08/03/2023
					CT2407	Active implantable	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	08/03/2023	08/03/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	08/03/2023	08/03/2023
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	08/03/2023	08/03/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	08/03/2023	08/03/2023
					CT2727	Active implant devices/ancillaries	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	08/03/2023	08/03/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	08/03/2023	08/03/2023
					CT2750	Implantable cardiac electrical stimulators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	08/03/2023	08/03/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	08/03/2023	08/03/2023
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	08/03/2023	08/03/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	08/03/2023	08/03/2023
					CT2985	Active	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	08/03/2023	08/03/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	08/03/2023	08/03/2023
					CT3001	Implantable metal	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	08/03/2023	08/03/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	08/03/2023	08/03/2023
					CT3002	Implantable synthetic polymer	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	08/03/2023	08/03/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	08/03/2023	08/03/2023
					CT3006	Implantable inorganic material	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	08/03/2023	08/03/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	08/03/2023	08/03/2023
					CT314	Electrically powered	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	08/03/2023	08/03/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	08/03/2023	08/03/2023
					CT321	Long-term surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	08/03/2023	08/03/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	08/03/2023	08/03/2023
					CT449	Generators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	08/03/2023	08/03/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	08/03/2023	08/03/2023
					CT981	Single use	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	08/03/2023	08/03/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	08/03/2023	08/03/2023
					CT983	Surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	08/03/2023	08/03/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	08/03/2023	08/03/2023
12593	2023	All other reportable incidents	47265	Dual-chamber implantable pacemaker, rate-responsive	CT141	Electrodes/Leads and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070101	Failure to Capture		
									A070102	High Capture Threshold		
									A070909	Oversensing		
									A0712	Pacing Problem		
							F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A070101	Failure to Capture		
									A070102	High Capture Threshold		
									A070909	Oversensing		
									A0712	Pacing Problem		
					CT145	Cardiovascular devices	E2403	No Clinical Signs, Symptoms or Conditions	A070101	Failure to Capture		
									A070102	High Capture Threshold		
									A070909	Oversensing		
									A0712	Pacing Problem		
							F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A070101	Failure to Capture		
									A070102	High Capture Threshold		
									A070909	Oversensing		
									A0712	Pacing Problem		
					CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070101	Failure to Capture		
									A070102	High Capture Threshold		
									A070909	Oversensing		
									A0712	Pacing Problem		
							F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A070101	Failure to Capture		
									A070102	High Capture Threshold		
									A070909	Oversensing		
									A0712	Pacing Problem		
					CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A070101	Failure to Capture		
									A070102	High Capture Threshold		
									A070909	Oversensing		
									A0712	Pacing Problem		
							F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A070101	Failure to Capture		

									A070102	High Capture Threshold		
									A070909	Oversensing		
									A0712	Pacing Problem		
					CT233	Surgical	E2403	No Clinical Signs, Symptoms or Conditions	A070101	Failure to Capture		
									A070102	High Capture Threshold		
									A070909	Oversensing		
									A0712	Pacing Problem		
							F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A070101	Failure to Capture		
									A070102	High Capture Threshold		
									A070909	Oversensing		
									A0712	Pacing Problem		
					CT2406	Implantable	E2403	No Clinical Signs, Symptoms or Conditions	A070101	Failure to Capture		
									A070102	High Capture Threshold		
									A070909	Oversensing		
									A0712	Pacing Problem		
							F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A070101	Failure to Capture		
									A070102	High Capture Threshold		
									A070909	Oversensing		
									A0712	Pacing Problem		
					CT2407	Active implantable	E2403	No Clinical Signs, Symptoms or Conditions	A070101	Failure to Capture		
									A070102	High Capture Threshold		
									A070909	Oversensing		
									A0712	Pacing Problem		
							F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A070101	Failure to Capture		
									A070102	High Capture Threshold		
									A070909	Oversensing		
									A0712	Pacing Problem		
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A070101	Failure to Capture		
									A070102	High Capture Threshold		
									A070909	Oversensing		
									A0712	Pacing Problem		
							F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A070101	Failure to Capture		
									A070102	High Capture Threshold		
									A070909	Oversensing		
									A0712	Pacing Problem		
					CT2727	Active implant devices/ancillaries	E2403	No Clinical Signs, Symptoms or Conditions	A070101	Failure to Capture		
									A070102	High Capture Threshold		
									A070909	Oversensing		
									A0712	Pacing Problem		
							F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A070101	Failure to Capture		
									A070102	High Capture Threshold		
									A070909	Oversensing		
									A0712	Pacing Problem		
					CT2750	Implantable cardiac electrical stimulators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070101	Failure to Capture		
									A070102	High Capture Threshold		
									A070909	Oversensing		
									A0712	Pacing Problem		
							F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A070101	Failure to Capture		
									A070102	High Capture Threshold		
									A070909	Oversensing		
									A0712	Pacing Problem		
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A070101	Failure to Capture		
									A070102	High Capture Threshold		
									A070909	Oversensing		
									A0712	Pacing Problem		
							F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A070101	Failure to Capture		
									A070102	High Capture Threshold		
									A070909	Oversensing		
									A0712	Pacing Problem		
					CT2985	Active	E2403	No Clinical Signs, Symptoms or Conditions	A070101	Failure to Capture		
									A070102	High Capture Threshold		
									A070909	Oversensing		
									A0712	Pacing Problem		
							F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A070101	Failure to Capture		
									A070102	High Capture Threshold		
									A070909	Oversensing		
									A0712	Pacing Problem		
					CT3001	Implantable metal	E2403	No Clinical Signs, Symptoms or Conditions	A070101	Failure to Capture		
									A070102	High Capture Threshold		
									A070909	Oversensing		
									A0712	Pacing Problem		
							F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A070101	Failure to Capture		
									A070102	High Capture Threshold		
									A070909	Oversensing		
									A0712	Pacing Problem		
					CT3002	Implantable synthetic polymer	E2403	No Clinical Signs, Symptoms or Conditions	A070101	Failure to Capture		
									A070102	High Capture Threshold		
									A070909	Oversensing		
									A0712	Pacing Problem		
							F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A070101	Failure to Capture		
									A070102	High Capture Threshold		
									A070909	Oversensing		
									A0712	Pacing Problem		
					CT3006	Implantable inorganic material	E2403	No Clinical Signs, Symptoms or Conditions	A070101	Failure to Capture		
									A070102	High Capture Threshold		
									A070909	Oversensing		
									A0712	Pacing Problem		
							F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A070101	Failure to Capture		

									A070102	High Capture Threshold		
									A070909	Oversensing		
									A0712	Pacing Problem		
					CT314	Electrically powered	E2403	No Clinical Signs, Symptoms or Conditions	A070101	Failure to Capture		
									A070102	High Capture Threshold		
									A070909	Oversensing		
									A0712	Pacing Problem		
							F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A070101	Failure to Capture		
									A070102	High Capture Threshold		
									A070909	Oversensing		
									A0712	Pacing Problem		
					CT321	Long-term surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A070101	Failure to Capture		
									A070102	High Capture Threshold		
									A070909	Oversensing		
									A0712	Pacing Problem		
							F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A070101	Failure to Capture		
									A070102	High Capture Threshold		
									A070909	Oversensing		
									A0712	Pacing Problem		
					CT449	Generators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070101	Failure to Capture		
									A070102	High Capture Threshold		
									A070909	Oversensing		
									A0712	Pacing Problem		
							F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A070101	Failure to Capture		
									A070102	High Capture Threshold		
									A070909	Oversensing		
									A0712	Pacing Problem		
					CT981	Single use	E2403	No Clinical Signs, Symptoms or Conditions	A070101	Failure to Capture		
									A070102	High Capture Threshold		
									A070909	Oversensing		
									A0712	Pacing Problem		
							F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A070101	Failure to Capture		
									A070102	High Capture Threshold		
									A070909	Oversensing		
									A0712	Pacing Problem		
					CT983	Surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A070101	Failure to Capture		
									A070102	High Capture Threshold		
									A070909	Oversensing		
									A0712	Pacing Problem		
							F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A070101	Failure to Capture		
									A070102	High Capture Threshold		
									A070909	Oversensing		
									A0712	Pacing Problem		
12594	2023	All other reportable incidents	47265	Dual-chamber implantable pacemaker, rate-responsive	CT141	Electrodes/Leads and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A0203	Defective Device		
							F26	No Health Consequences or Impact	A0203	Defective Device		
					CT145	Cardiovascular devices	E2403	No Clinical Signs, Symptoms or Conditions	A0203	Defective Device		
							F26	No Health Consequences or Impact	A0203	Defective Device		
					CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A0203	Defective Device		
							F26	No Health Consequences or Impact	A0203	Defective Device		
					CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A0203	Defective Device		
							F26	No Health Consequences or Impact	A0203	Defective Device		
					CT233	Surgical	E2403	No Clinical Signs, Symptoms or Conditions	A0203	Defective Device		
							F26	No Health Consequences or Impact	A0203	Defective Device		
					CT2406	Implantable	E2403	No Clinical Signs, Symptoms or Conditions	A0203	Defective Device		
							F26	No Health Consequences or Impact	A0203	Defective Device		
					CT2407	Active implantable	E2403	No Clinical Signs, Symptoms or Conditions	A0203	Defective Device		
							F26	No Health Consequences or Impact	A0203	Defective Device		
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A0203	Defective Device		
							F26	No Health Consequences or Impact	A0203	Defective Device		
					CT2727	Active implant devices/ancillaries	E2403	No Clinical Signs, Symptoms or Conditions	A0203	Defective Device		
							F26	No Health Consequences or Impact	A0203	Defective Device		
					CT2750	Implantable cardiac electrical stimulators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A0203	Defective Device		
							F26	No Health Consequences or Impact	A0203	Defective Device		
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A0203	Defective Device		
							F26	No Health Consequences or Impact	A0203	Defective Device		
					CT2985	Active	E2403	No Clinical Signs, Symptoms or Conditions	A0203	Defective Device		
							F26	No Health Consequences or Impact	A0203	Defective Device		
					CT3001	Implantable metal	E2403	No Clinical Signs, Symptoms or Conditions	A0203	Defective Device		
							F26	No Health Consequences or Impact	A0203	Defective Device		
					CT3002	Implantable synthetic polymer	E2403	No Clinical Signs, Symptoms or Conditions	A0203	Defective Device		
							F26	No Health Consequences or Impact	A0203	Defective Device		
					CT3006	Implantable inorganic material	E2403	No Clinical Signs, Symptoms or Conditions	A0203	Defective Device		
							F26	No Health Consequences or Impact	A0203	Defective Device		
					CT314	Electrically powered	E2403	No Clinical Signs, Symptoms or Conditions	A0203	Defective Device		
							F26	No Health Consequences or Impact	A0203	Defective Device		
					CT321	Long-term surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A0203	Defective Device		
							F26	No Health Consequences or Impact	A0203	Defective Device		
					CT449	Generators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A0203	Defective Device		
							F26	No Health Consequences or Impact	A0203	Defective Device		
					CT981	Single use	E2403	No Clinical Signs, Symptoms or Conditions	A0203	Defective Device		
							F26	No Health Consequences or Impact	A0203	Defective Device		
					CT983	Surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A0203	Defective Device		
							F26	No Health Consequences or Impact	A0203	Defective Device		
12595	2023	All other reportable incidents	36241	Pacemaker/defibrillator lead	CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A072202	Low Impedance		
									A090801	Signal Artifact/Noise		
							F26	No Health Consequences or Impact	A072202	Low Impedance		
									A090801	Signal Artifact/Noise		
12596	2023	All other reportable incidents	47265	Dual-chamber implantable pacemaker, rate-	CT141	Electrodes/Leads and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem		

				responsive					A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
					CT145	Cardiovascular devices	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
					CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
					CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
					CT233	Surgical	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
					CT2406	Implantable	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
					CT2407	Active implantable	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
					CT2727	Active implant devices/ancillaries	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
					CT2750	Implantable cardiac electrical stimulators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
					CT2985	Active	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
					CT3001	Implantable metal	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
					CT3002	Implantable synthetic polymer	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
					CT3006	Implantable inorganic material	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
					CT314	Electrically powered	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
					CT321	Long-term surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
					CT449	Generators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
					CT981	Single use	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
					CT983	Surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
12597	2023	All other reportable incidents	47265	Dual-chamber implantable pacemaker, rate-responsive	CT141	Electrodes/Leads and associated devices	E011903	Syncope/Fainting	A070102	High Capture Threshold		
									A070909	Oversensing		
									A0712	Pacing Problem		
									A090801	Signal Artifact/Noise		
							F26	No Health Consequences or Impact	A070102	High Capture Threshold		
									A070909	Oversensing		
									A0712	Pacing Problem		
									A090801	Signal Artifact/Noise		
					CT145	Cardiovascular devices	E011903	Syncope/Fainting	A070102	High Capture Threshold		

									A070909	Oversensing		
									A0712	Pacing Problem		
									A090801	Signal Artifact/Noise		
								F26	No Health Consequences or Impact	A070102	High Capture Threshold		
										A070909	Oversensing		
										A0712	Pacing Problem		
										A090801	Signal Artifact/Noise		
						CT1684	Cardiac pacemakers and associated devices	E011903	Syncope/Fainting	A070102	High Capture Threshold		
										A070909	Oversensing		
										A0712	Pacing Problem		
										A090801	Signal Artifact/Noise		
								F26	No Health Consequences or Impact	A070102	High Capture Threshold		
										A070909	Oversensing		
										A0712	Pacing Problem		
										A090801	Signal Artifact/Noise		
						CT1685	Cardiac pacemakers	E011903	Syncope/Fainting	A070102	High Capture Threshold		
										A070909	Oversensing		
										A0712	Pacing Problem		
										A090801	Signal Artifact/Noise		
								F26	No Health Consequences or Impact	A070102	High Capture Threshold		
										A070909	Oversensing		
										A0712	Pacing Problem		
										A090801	Signal Artifact/Noise		
						CT233	Surgical	E011903	Syncope/Fainting	A070102	High Capture Threshold		
										A070909	Oversensing		
										A0712	Pacing Problem		
										A090801	Signal Artifact/Noise		
								F26	No Health Consequences or Impact	A070102	High Capture Threshold		
										A070909	Oversensing		
										A0712	Pacing Problem		
										A090801	Signal Artifact/Noise		
						CT2406	Implantable	E011903	Syncope/Fainting	A070102	High Capture Threshold		
										A070909	Oversensing		
										A0712	Pacing Problem		
										A090801	Signal Artifact/Noise		
								F26	No Health Consequences or Impact	A070102	High Capture Threshold		
										A070909	Oversensing		
										A0712	Pacing Problem		
										A090801	Signal Artifact/Noise		
						CT2407	Active implantable	E011903	Syncope/Fainting	A070102	High Capture Threshold		
										A070909	Oversensing		
										A0712	Pacing Problem		
										A090801	Signal Artifact/Noise		
								F26	No Health Consequences or Impact	A070102	High Capture Threshold		
										A070909	Oversensing		
										A0712	Pacing Problem		
										A090801	Signal Artifact/Noise		
						CT257	Implantable cardiac pacemakers	E011903	Syncope/Fainting	A070102	High Capture Threshold		
										A070909	Oversensing		
										A0712	Pacing Problem		
										A090801	Signal Artifact/Noise		
								F26	No Health Consequences or Impact	A070102	High Capture Threshold		
										A070909	Oversensing		
										A0712	Pacing Problem		
										A090801	Signal Artifact/Noise		
						CT2727	Active implant devices/ancillaries	E011903	Syncope/Fainting	A070102	High Capture Threshold		
										A070909	Oversensing		
										A0712	Pacing Problem		
										A090801	Signal Artifact/Noise		
								F26	No Health Consequences or Impact	A070102	High Capture Threshold		
										A070909	Oversensing		
										A0712	Pacing Problem		
										A090801	Signal Artifact/Noise		
						CT2750	Implantable cardiac electrical stimulators and associated devices	E011903	Syncope/Fainting	A070102	High Capture Threshold		
										A070909	Oversensing		
										A0712	Pacing Problem		
										A090801	Signal Artifact/Noise		
								F26	No Health Consequences or Impact	A070102	High Capture Threshold		
										A070909	Oversensing		
										A0712	Pacing Problem		
										A090801	Signal Artifact/Noise		
						CT2866	Implantable cardiac electrical stimulators	E011903	Syncope/Fainting	A070102	High Capture Threshold		
										A070909	Oversensing		
										A0712	Pacing Problem		
										A090801	Signal Artifact/Noise		
								F26	No Health Consequences or Impact	A070102	High Capture Threshold		
										A070909	Oversensing		
										A0712	Pacing Problem		
										A090801	Signal Artifact/Noise		
						CT2985	Active	E011903	Syncope/Fainting	A070102	High Capture Threshold		
										A070909	Oversensing		
										A0712	Pacing Problem		
										A090801	Signal Artifact/Noise		
								F26	No Health Consequences or Impact	A070102	High Capture Threshold		
										A070909	Oversensing		
										A0712	Pacing Problem		
										A090801	Signal Artifact/Noise		
						CT3001	Implantable metal	E011903	Syncope/Fainting	A070102	High Capture Threshold		

									A070909	Oversensing		
									A0712	Pacing Problem		
									A090801	Signal Artifact/Noise		
							F26	No Health Consequences or Impact	A070102	High Capture Threshold		
									A070909	Oversensing		
									A0712	Pacing Problem		
									A090801	Signal Artifact/Noise		
					CT3002	Implantable synthetic polymer	E011903	Syncope/Fainting	A070102	High Capture Threshold		
									A070909	Oversensing		
									A0712	Pacing Problem		
									A090801	Signal Artifact/Noise		
							F26	No Health Consequences or Impact	A070102	High Capture Threshold		
									A070909	Oversensing		
									A0712	Pacing Problem		
									A090801	Signal Artifact/Noise		
					CT3006	Implantable inorganic material	E011903	Syncope/Fainting	A070102	High Capture Threshold		
									A070909	Oversensing		
									A0712	Pacing Problem		
									A090801	Signal Artifact/Noise		
							F26	No Health Consequences or Impact	A070102	High Capture Threshold		
									A070909	Oversensing		
									A0712	Pacing Problem		
									A090801	Signal Artifact/Noise		
					CT314	Electrically powered	E011903	Syncope/Fainting	A070102	High Capture Threshold		
									A070909	Oversensing		
									A0712	Pacing Problem		
									A090801	Signal Artifact/Noise		
							F26	No Health Consequences or Impact	A070102	High Capture Threshold		
									A070909	Oversensing		
									A0712	Pacing Problem		
									A090801	Signal Artifact/Noise		
					CT321	Long-term surgical invasive	E011903	Syncope/Fainting	A070102	High Capture Threshold		
									A070909	Oversensing		
									A0712	Pacing Problem		
									A090801	Signal Artifact/Noise		
							F26	No Health Consequences or Impact	A070102	High Capture Threshold		
									A070909	Oversensing		
									A0712	Pacing Problem		
									A090801	Signal Artifact/Noise		
					CT449	Generators and associated devices	E011903	Syncope/Fainting	A070102	High Capture Threshold		
									A070909	Oversensing		
									A0712	Pacing Problem		
									A090801	Signal Artifact/Noise		
							F26	No Health Consequences or Impact	A070102	High Capture Threshold		
									A070909	Oversensing		
									A0712	Pacing Problem		
									A090801	Signal Artifact/Noise		
					CT981	Single use	E011903	Syncope/Fainting	A070102	High Capture Threshold		
									A070909	Oversensing		
									A0712	Pacing Problem		
									A090801	Signal Artifact/Noise		
							F26	No Health Consequences or Impact	A070102	High Capture Threshold		
									A070909	Oversensing		
									A0712	Pacing Problem		
									A090801	Signal Artifact/Noise		
					CT983	Surgical invasive	E011903	Syncope/Fainting	A070102	High Capture Threshold		
									A070909	Oversensing		
									A0712	Pacing Problem		
									A090801	Signal Artifact/Noise		
							F26	No Health Consequences or Impact	A070102	High Capture Threshold		
									A070909	Oversensing		
									A0712	Pacing Problem		
									A090801	Signal Artifact/Noise		
12598	2023	All other reportable incidents	47263	Cardiac resynchronization therapy implantable pacemaker	CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A0908	Incorrect, Inadequate or Imprecise Result or Readings	27/02/2023	27/02/2023
							F08	Hospitalization or Prolonged Hospitalization	A0908	Incorrect, Inadequate or Imprecise Result or Readings	27/02/2023	27/02/2023
							F1905	Device Revision or Replacement	A0908	Incorrect, Inadequate or Imprecise Result or Readings	27/02/2023	27/02/2023
					CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A0908	Incorrect, Inadequate or Imprecise Result or Readings	27/02/2023	27/02/2023
							F08	Hospitalization or Prolonged Hospitalization	A0908	Incorrect, Inadequate or Imprecise Result or Readings	27/02/2023	27/02/2023
							F1905	Device Revision or Replacement	A0908	Incorrect, Inadequate or Imprecise Result or Readings	27/02/2023	27/02/2023
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A0908	Incorrect, Inadequate or Imprecise Result or Readings	27/02/2023	27/02/2023
							F08	Hospitalization or Prolonged Hospitalization	A0908	Incorrect, Inadequate or Imprecise Result or Readings	27/02/2023	27/02/2023
							F1905	Device Revision or Replacement	A0908	Incorrect, Inadequate or Imprecise Result or Readings	27/02/2023	27/02/2023
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A0908	Incorrect, Inadequate or Imprecise Result or Readings	27/02/2023	27/02/2023
							F08	Hospitalization or Prolonged Hospitalization	A0908	Incorrect, Inadequate or Imprecise Result or Readings	27/02/2023	27/02/2023
							F1905	Device Revision or Replacement	A0908	Incorrect, Inadequate or Imprecise Result or Readings	27/02/2023	27/02/2023
12599	2023	All other reportable incidents	47265	Dual-chamber implantable pacemaker, rate-responsive	CT141	Electrodes/Leads and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070909	Oversensing		
									A090801	Signal Artifact/Noise		
							F26	No Health Consequences or Impact	A070909	Oversensing		
									A090801	Signal Artifact/Noise		

					CT145	Cardiovascular devices	E2403	No Clinical Signs, Symptoms or Conditions	A070909	Oversensing		
									A090801	Signal Artifact/Noise		
							F26	No Health Consequences or Impact	A070909	Oversensing		
									A090801	Signal Artifact/Noise		
					CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070909	Oversensing		
									A090801	Signal Artifact/Noise		
							F26	No Health Consequences or Impact	A070909	Oversensing		
									A090801	Signal Artifact/Noise		
					CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A070909	Oversensing		
									A090801	Signal Artifact/Noise		
							F26	No Health Consequences or Impact	A070909	Oversensing		
									A090801	Signal Artifact/Noise		
					CT233	Surgical	E2403	No Clinical Signs, Symptoms or Conditions	A070909	Oversensing		
									A090801	Signal Artifact/Noise		
							F26	No Health Consequences or Impact	A070909	Oversensing		
									A090801	Signal Artifact/Noise		
					CT2406	Implantable	E2403	No Clinical Signs, Symptoms or Conditions	A070909	Oversensing		
									A090801	Signal Artifact/Noise		
							F26	No Health Consequences or Impact	A070909	Oversensing		
									A090801	Signal Artifact/Noise		
					CT2407	Active implantable	E2403	No Clinical Signs, Symptoms or Conditions	A070909	Oversensing		
									A090801	Signal Artifact/Noise		
							F26	No Health Consequences or Impact	A070909	Oversensing		
									A090801	Signal Artifact/Noise		
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A070909	Oversensing		
									A090801	Signal Artifact/Noise		
							F26	No Health Consequences or Impact	A070909	Oversensing		
									A090801	Signal Artifact/Noise		
					CT2727	Active implant devices/ancillaries	E2403	No Clinical Signs, Symptoms or Conditions	A070909	Oversensing		
									A090801	Signal Artifact/Noise		
							F26	No Health Consequences or Impact	A070909	Oversensing		
									A090801	Signal Artifact/Noise		
					CT2750	Implantable cardiac electrical stimulators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070909	Oversensing		
									A090801	Signal Artifact/Noise		
							F26	No Health Consequences or Impact	A070909	Oversensing		
									A090801	Signal Artifact/Noise		
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A070909	Oversensing		
									A090801	Signal Artifact/Noise		
							F26	No Health Consequences or Impact	A070909	Oversensing		
									A090801	Signal Artifact/Noise		
					CT2985	Active	E2403	No Clinical Signs, Symptoms or Conditions	A070909	Oversensing		
									A090801	Signal Artifact/Noise		
							F26	No Health Consequences or Impact	A070909	Oversensing		
									A090801	Signal Artifact/Noise		
					CT3001	Implantable metal	E2403	No Clinical Signs, Symptoms or Conditions	A070909	Oversensing		
									A090801	Signal Artifact/Noise		
							F26	No Health Consequences or Impact	A070909	Oversensing		
									A090801	Signal Artifact/Noise		
					CT3002	Implantable synthetic polymer	E2403	No Clinical Signs, Symptoms or Conditions	A070909	Oversensing		
									A090801	Signal Artifact/Noise		
							F26	No Health Consequences or Impact	A070909	Oversensing		
									A090801	Signal Artifact/Noise		
					CT3006	Implantable inorganic material	E2403	No Clinical Signs, Symptoms or Conditions	A070909	Oversensing		
									A090801	Signal Artifact/Noise		
							F26	No Health Consequences or Impact	A070909	Oversensing		
									A090801	Signal Artifact/Noise		
					CT314	Electrically powered	E2403	No Clinical Signs, Symptoms or Conditions	A070909	Oversensing		
									A090801	Signal Artifact/Noise		
							F26	No Health Consequences or Impact	A070909	Oversensing		
									A090801	Signal Artifact/Noise		
					CT321	Long-term surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A070909	Oversensing		
									A090801	Signal Artifact/Noise		
							F26	No Health Consequences or Impact	A070909	Oversensing		
									A090801	Signal Artifact/Noise		
					CT449	Generators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070909	Oversensing		
									A090801	Signal Artifact/Noise		
							F26	No Health Consequences or Impact	A070909	Oversensing		
									A090801	Signal Artifact/Noise		
					CT981	Single use	E2403	No Clinical Signs, Symptoms or Conditions	A070909	Oversensing		
									A090801	Signal Artifact/Noise		
							F26	No Health Consequences or Impact	A070909	Oversensing		
									A090801	Signal Artifact/Noise		
					CT983	Surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A070909	Oversensing		
									A090801	Signal Artifact/Noise		
							F26	No Health Consequences or Impact	A070909	Oversensing		
									A090801	Signal Artifact/Noise		
12600	2023	All other reportable incidents	47265	Dual-chamber implantable pacemaker, rate-responsive	CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A0908	Incorrect, Inadequate or Imprecise Result or Readings	01/03/2023	01/03/2023
							F08	Hospitalization or Prolonged Hospitalization	A0908	Incorrect, Inadequate or Imprecise Result or Readings	01/03/2023	01/03/2023
							F1905	Device Revision or Replacement	A0908	Incorrect, Inadequate or Imprecise Result or Readings	01/03/2023	01/03/2023
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A0908	Incorrect, Inadequate or Imprecise Result or Readings	01/03/2023	01/03/2023
							F08	Hospitalization or Prolonged Hospitalization	A0908	Incorrect, Inadequate or Imprecise Result or Readings	01/03/2023	01/03/2023
							F1905	Device Revision or Replacement	A0908	Incorrect, Inadequate or Imprecise Result or Readings	01/03/2023	01/03/2023

					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A0908	Incorrect, Inadequate or Imprecise Result or Readings	01/03/2023	01/03/2023
							F08	Hospitalization or Prolonged Hospitalization	A0908	Incorrect, Inadequate or Imprecise Result or Readings	01/03/2023	01/03/2023
							F1905	Device Revision or Replacement	A0908	Incorrect, Inadequate or Imprecise Result or Readings	01/03/2023	01/03/2023
12601	2023	All other reportable incidents	35852	Single-chamber implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	02/01/2023	02/01/2023
									A0908	Incorrect, Inadequate or Imprecise Result or Readings	02/01/2023	02/01/2023
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	02/01/2023	02/01/2023
									A0908	Incorrect, Inadequate or Imprecise Result or Readings	02/01/2023	02/01/2023
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	02/01/2023	02/01/2023
									A0908	Incorrect, Inadequate or Imprecise Result or Readings	02/01/2023	02/01/2023
12602	2023		47265	Dual-chamber implantable pacemaker, rate-responsive	CT141	Electrodes/Leads and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	03/02/2023	03/02/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	03/02/2023	03/02/2023
					CT145	Cardiovascular devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	03/02/2023	03/02/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	03/02/2023	03/02/2023
					CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	03/02/2023	03/02/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	03/02/2023	03/02/2023
					CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	03/02/2023	03/02/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	03/02/2023	03/02/2023
					CT233	Surgical	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	03/02/2023	03/02/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	03/02/2023	03/02/2023
					CT2406	Implantable	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	03/02/2023	03/02/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	03/02/2023	03/02/2023
					CT2407	Active implantable	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	03/02/2023	03/02/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	03/02/2023	03/02/2023
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	03/02/2023	03/02/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	03/02/2023	03/02/2023
					CT2727	Active implant devices/ancillaries	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	03/02/2023	03/02/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	03/02/2023	03/02/2023
					CT2750	Implantable cardiac electrical stimulators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	03/02/2023	03/02/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	03/02/2023	03/02/2023
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	03/02/2023	03/02/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	03/02/2023	03/02/2023
					CT2985	Active	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	03/02/2023	03/02/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	03/02/2023	03/02/2023
					CT3001	Implantable metal	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	03/02/2023	03/02/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	03/02/2023	03/02/2023
					CT3002	Implantable synthetic polymer	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	03/02/2023	03/02/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	03/02/2023	03/02/2023
					CT3006	Implantable inorganic material	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	03/02/2023	03/02/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	03/02/2023	03/02/2023
					CT314	Electrically powered	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	03/02/2023	03/02/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	03/02/2023	03/02/2023
					CT321	Long-term surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	03/02/2023	03/02/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	03/02/2023	03/02/2023
					CT449	Generators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	03/02/2023	03/02/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	03/02/2023	03/02/2023
					CT981	Single use	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	03/02/2023	03/02/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	03/02/2023	03/02/2023
					CT983	Surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	03/02/2023	03/02/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	03/02/2023	03/02/2023
12603	2023		47265	Dual-chamber implantable pacemaker, rate-responsive	CT141	Electrodes/Leads and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	16/03/2023	16/03/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	16/03/2023	16/03/2023
					CT145	Cardiovascular devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	16/03/2023	16/03/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	16/03/2023	16/03/2023
					CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	16/03/2023	16/03/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	16/03/2023	16/03/2023
					CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	16/03/2023	16/03/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	16/03/2023	16/03/2023
					CT233	Surgical	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	16/03/2023	16/03/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	16/03/2023	16/03/2023
					CT2406	Implantable	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	16/03/2023	16/03/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	16/03/2023	16/03/2023
					CT2407	Active implantable	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	16/03/2023	16/03/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	16/03/2023	16/03/2023
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	16/03/2023	16/03/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	16/03/2023	16/03/2023
					CT2727	Active implant devices/ancillaries	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	16/03/2023	16/03/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	16/03/2023	16/03/2023
					CT2750	Implantable cardiac electrical stimulators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	16/03/2023	16/03/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	16/03/2023	16/03/2023
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	16/03/2023	16/03/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	16/03/2023	16/03/2023
					CT2985	Active	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	16/03/2023	16/03/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	16/03/2023	16/03/2023
					CT3001	Implantable metal	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	16/03/2023	16/03/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	16/03/2023	16/03/2023
					CT3002	Implantable synthetic polymer	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	16/03/2023	16/03/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	16/03/2023	16/03/2023
					CT3006	Implantable inorganic material	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	16/03/2023	16/03/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	16/03/2023	16/03/2023
					CT314	Electrically powered	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	16/03/2023	16/03/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	16/03/2023	16/03/2023
					CT321	Long-term surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	16/03/2023	16/03/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	16/03/2023	16/03/2023

					CT449	Generators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	16/03/2023	16/03/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	16/03/2023	16/03/2023
					CT981	Single use	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	16/03/2023	16/03/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	16/03/2023	16/03/2023
					CT983	Surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	16/03/2023	16/03/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	16/03/2023	16/03/2023
12604	2023	All other reportable incidents	17882	Physiologic-monitoring defibrillation system	CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A150105	Key or Button Unresponsive/not Working		
							F27	Problem Identified During Non-Clinical Procedure	A150105	Key or Button Unresponsive/not Working		
12605	2023	All other reportable incidents	35852	Single-chamber implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E2104	Electric Shock	A0709	Device Sensing Problem	01/12/2022	01/12/2022
									A070909	Oversensing	01/12/2022	01/12/2022
									A071302	Inappropriate/Inadequate Shock/Stimulation	01/12/2022	01/12/2022
									A090801	Signal Artifact/Noise	01/12/2022	01/12/2022
							F08	Hospitalization or Prolonged Hospitalization	A0709	Device Sensing Problem	01/12/2022	01/12/2022
									A070909	Oversensing	01/12/2022	01/12/2022
									A071302	Inappropriate/Inadequate Shock/Stimulation	01/12/2022	01/12/2022
									A090801	Signal Artifact/Noise	01/12/2022	01/12/2022
							F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A0709	Device Sensing Problem	01/12/2022	01/12/2022
									A070909	Oversensing	01/12/2022	01/12/2022
									A071302	Inappropriate/Inadequate Shock/Stimulation	01/12/2022	01/12/2022
									A090801	Signal Artifact/Noise	01/12/2022	01/12/2022
							F1903	Device Explantation	A0709	Device Sensing Problem	01/12/2022	01/12/2022
									A070909	Oversensing	01/12/2022	01/12/2022
									A071302	Inappropriate/Inadequate Shock/Stimulation	01/12/2022	01/12/2022
									A090801	Signal Artifact/Noise	01/12/2022	01/12/2022
							F2203	Imaging Required	A0709	Device Sensing Problem	01/12/2022	01/12/2022
									A070909	Oversensing	01/12/2022	01/12/2022
									A071302	Inappropriate/Inadequate Shock/Stimulation	01/12/2022	01/12/2022
									A090801	Signal Artifact/Noise	01/12/2022	01/12/2022
							F23	Unexpected Medical Intervention	A0709	Device Sensing Problem	01/12/2022	01/12/2022
									A070909	Oversensing	01/12/2022	01/12/2022
									A071302	Inappropriate/Inadequate Shock/Stimulation	01/12/2022	01/12/2022
									A090801	Signal Artifact/Noise	01/12/2022	01/12/2022
12606	2023	All other reportable incidents	35852	Single-chamber implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
									A160601	Premature Elective Replacement Indicator		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
									A160601	Premature Elective Replacement Indicator		
12607	2023	All other reportable incidents	47265	Dual-chamber implantable pacemaker, rate-responsive	CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	07/12/2022	07/12/2022
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	07/12/2022	07/12/2022
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	07/12/2022	07/12/2022
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	07/12/2022	07/12/2022
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	07/12/2022	07/12/2022
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	07/12/2022	07/12/2022
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	07/12/2022	07/12/2022
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	07/12/2022	07/12/2022
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	07/12/2022	07/12/2022
12608	2023	All other reportable incidents	17882	Physiologic-monitoring defibrillation system	CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A07	Electrical /Electronic Property Problem		
							F27	Problem Identified During Non-Clinical Procedure	A07	Electrical /Electronic Property Problem		
12609	2023	All other reportable incidents	17882	Physiologic-monitoring defibrillation system	CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A07	Electrical /Electronic Property Problem		
							F27	Problem Identified During Non-Clinical Procedure	A07	Electrical /Electronic Property Problem		
12610	2023	All other reportable incidents	17882	Physiologic-monitoring defibrillation system	CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A07	Electrical /Electronic Property Problem		
							F27	Problem Identified During Non-Clinical Procedure	A07	Electrical /Electronic Property Problem		
12611	2023	All other reportable incidents	47263	Cardiac resynchronization therapy implantable pacemaker	CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A0908	Incorrect, Inadequate or Imprecise Result or Readings	08/12/2022	08/12/2022
							F08	Hospitalization or Prolonged Hospitalization	A0908	Incorrect, Inadequate or Imprecise Result or Readings	08/12/2022	08/12/2022
							F1905	Device Revision or Replacement	A0908	Incorrect, Inadequate or Imprecise Result or Readings	08/12/2022	08/12/2022
					CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A0908	Incorrect, Inadequate or Imprecise Result or Readings	08/12/2022	08/12/2022
							F08	Hospitalization or Prolonged Hospitalization	A0908	Incorrect, Inadequate or Imprecise Result or Readings	08/12/2022	08/12/2022
							F1905	Device Revision or Replacement	A0908	Incorrect, Inadequate or Imprecise Result or Readings	08/12/2022	08/12/2022
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A0908	Incorrect, Inadequate or Imprecise Result or Readings	08/12/2022	08/12/2022
							F08	Hospitalization or Prolonged Hospitalization	A0908	Incorrect, Inadequate or Imprecise Result or Readings	08/12/2022	08/12/2022
							F1905	Device Revision or Replacement	A0908	Incorrect, Inadequate or Imprecise Result or Readings	08/12/2022	08/12/2022
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A0908	Incorrect, Inadequate or Imprecise Result or Readings	08/12/2022	08/12/2022
							F08	Hospitalization or Prolonged Hospitalization	A0908	Incorrect, Inadequate or Imprecise Result or Readings	08/12/2022	08/12/2022
							F1905	Device Revision or Replacement	A0908	Incorrect, Inadequate or Imprecise Result or Readings	08/12/2022	08/12/2022
12612	2023	All other reportable incidents	47265	Dual-chamber implantable pacemaker, rate-responsive	CT141	Electrodes/Leads and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A072201	High Impedance		
							F26	No Health Consequences or Impact	A072201	High Impedance		
					CT145	Cardiovascular devices	E2403	No Clinical Signs, Symptoms or Conditions	A072201	High Impedance		
							F26	No Health Consequences or Impact	A072201	High Impedance		
					CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A072201	High Impedance		
							F26	No Health Consequences or Impact	A072201	High Impedance		
					CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A072201	High Impedance		
							F26	No Health Consequences or Impact	A072201	High Impedance		
					CT233	Surgical	E2403	No Clinical Signs, Symptoms or Conditions	A072201	High Impedance		
							F26	No Health Consequences or Impact	A072201	High Impedance		
					CT2406	Implantable	E2403	No Clinical Signs, Symptoms or Conditions	A072201	High Impedance		
							F26	No Health Consequences or Impact	A072201	High Impedance		
					CT2407	Active implantable	E2403	No Clinical Signs, Symptoms or Conditions	A072201	High Impedance		
							F26	No Health Consequences or Impact	A072201	High Impedance		
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A072201	High Impedance		
							F26	No Health Consequences or Impact	A072201	High Impedance		

					CT2727	Active implant devices/ancillaries	E2403	No Clinical Signs, Symptoms or Conditions	A072201	High Impedance		
							F26	No Health Consequences or Impact	A072201	High Impedance		
					CT2750	Implantable cardiac electrical stimulators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A072201	High Impedance		
							F26	No Health Consequences or Impact	A072201	High Impedance		
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A072201	High Impedance		
							F26	No Health Consequences or Impact	A072201	High Impedance		
					CT2985	Active	E2403	No Clinical Signs, Symptoms or Conditions	A072201	High Impedance		
							F26	No Health Consequences or Impact	A072201	High Impedance		
					CT3001	Implantable metal	E2403	No Clinical Signs, Symptoms or Conditions	A072201	High Impedance		
							F26	No Health Consequences or Impact	A072201	High Impedance		
					CT3002	Implantable synthetic polymer	E2403	No Clinical Signs, Symptoms or Conditions	A072201	High Impedance		
							F26	No Health Consequences or Impact	A072201	High Impedance		
					CT3006	Implantable inorganic material	E2403	No Clinical Signs, Symptoms or Conditions	A072201	High Impedance		
							F26	No Health Consequences or Impact	A072201	High Impedance		
					CT314	Electrically powered	E2403	No Clinical Signs, Symptoms or Conditions	A072201	High Impedance		
							F26	No Health Consequences or Impact	A072201	High Impedance		
					CT321	Long-term surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A072201	High Impedance		
							F26	No Health Consequences or Impact	A072201	High Impedance		
					CT449	Generators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A072201	High Impedance		
							F26	No Health Consequences or Impact	A072201	High Impedance		
					CT981	Single use	E2403	No Clinical Signs, Symptoms or Conditions	A072201	High Impedance		
							F26	No Health Consequences or Impact	A072201	High Impedance		
					CT983	Surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A072201	High Impedance		
							F26	No Health Consequences or Impact	A072201	High Impedance		
12613	2023	All other reportable incidents	47265	Dual-chamber implantable pacemaker, rate-responsive	CT141	Electrodes/Leads and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	27/01/2023	27/01/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	27/01/2023	27/01/2023
					CT145	Cardiovascular devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	27/01/2023	27/01/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	27/01/2023	27/01/2023
					CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	27/01/2023	27/01/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	27/01/2023	27/01/2023
					CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	27/01/2023	27/01/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	27/01/2023	27/01/2023
					CT233	Surgical	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	27/01/2023	27/01/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	27/01/2023	27/01/2023
					CT2406	Implantable	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	27/01/2023	27/01/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	27/01/2023	27/01/2023
					CT2407	Active implantable	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	27/01/2023	27/01/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	27/01/2023	27/01/2023
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	27/01/2023	27/01/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	27/01/2023	27/01/2023
					CT2727	Active implant devices/ancillaries	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	27/01/2023	27/01/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	27/01/2023	27/01/2023
					CT2750	Implantable cardiac electrical stimulators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	27/01/2023	27/01/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	27/01/2023	27/01/2023
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	27/01/2023	27/01/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	27/01/2023	27/01/2023
					CT2985	Active	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	27/01/2023	27/01/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	27/01/2023	27/01/2023
					CT3001	Implantable metal	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	27/01/2023	27/01/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	27/01/2023	27/01/2023
					CT3002	Implantable synthetic polymer	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	27/01/2023	27/01/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	27/01/2023	27/01/2023
					CT3006	Implantable inorganic material	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	27/01/2023	27/01/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	27/01/2023	27/01/2023
					CT314	Electrically powered	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	27/01/2023	27/01/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	27/01/2023	27/01/2023
					CT321	Long-term surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	27/01/2023	27/01/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	27/01/2023	27/01/2023
					CT449	Generators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	27/01/2023	27/01/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	27/01/2023	27/01/2023
					CT981	Single use	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	27/01/2023	27/01/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	27/01/2023	27/01/2023
					CT983	Surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	27/01/2023	27/01/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	27/01/2023	27/01/2023
12614	2023		47265	Dual-chamber implantable pacemaker, rate-responsive	CT141	Electrodes/Leads and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	27/01/2023	27/01/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	27/01/2023	27/01/2023
					CT145	Cardiovascular devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	27/01/2023	27/01/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	27/01/2023	27/01/2023
					CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	27/01/2023	27/01/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	27/01/2023	27/01/2023
					CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	27/01/2023	27/01/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	27/01/2023	27/01/2023
					CT233	Surgical	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	27/01/2023	27/01/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	27/01/2023	27/01/2023
					CT2406	Implantable	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	27/01/2023	27/01/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	27/01/2023	27/01/2023
					CT2407	Active implantable	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	27/01/2023	27/01/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	27/01/2023	27/01/2023
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	27/01/2023	27/01/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	27/01/2023	27/01/2023
					CT2727	Active implant devices/ancillaries	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	27/01/2023	27/01/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	27/01/2023	27/01/2023
					CT2750	Implantable cardiac electrical stimulators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	27/01/2023	27/01/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	27/01/2023	27/01/2023
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	27/01/2023	27/01/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	27/01/2023	27/01/2023
					CT2985	Active	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	27/01/2023	27/01/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	27/01/2023	27/01/2023

					CT3001	Implantable metal	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	27/01/2023	27/01/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	27/01/2023	27/01/2023
					CT3002	Implantable synthetic polymer	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	27/01/2023	27/01/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	27/01/2023	27/01/2023
					CT3006	Implantable inorganic material	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	27/01/2023	27/01/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	27/01/2023	27/01/2023
					CT314	Electrically powered	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	27/01/2023	27/01/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	27/01/2023	27/01/2023
					CT321	Long-term surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	27/01/2023	27/01/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	27/01/2023	27/01/2023
					CT449	Generators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	27/01/2023	27/01/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	27/01/2023	27/01/2023
					CT981	Single use	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	27/01/2023	27/01/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	27/01/2023	27/01/2023
					CT983	Surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	27/01/2023	27/01/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	27/01/2023	27/01/2023
12615	2023	All other reportable incidents	63188	Left-ventricular wireless pacing system sensor/transmitter	CT145	Cardiovascular devices	E2402	Appropriate Clinical Signs, Symptoms and Conditions Term/Code not Available	A070801	Complete Loss of Power	28/07/2022	28/07/2022
							F26	No Health Consequences or Impact	A070801	Complete Loss of Power	28/07/2022	28/07/2022
					CT1577	Wireless monitoring system transmitters	E2402	Appropriate Clinical Signs, Symptoms and Conditions Term/Code not Available	A070801	Complete Loss of Power	28/07/2022	28/07/2022
							F26	No Health Consequences or Impact	A070801	Complete Loss of Power	28/07/2022	28/07/2022
					CT1684	Cardiac pacemakers and associated devices	E2402	Appropriate Clinical Signs, Symptoms and Conditions Term/Code not Available	A070801	Complete Loss of Power	28/07/2022	28/07/2022
							F26	No Health Consequences or Impact	A070801	Complete Loss of Power	28/07/2022	28/07/2022
					CT2727	Active implant devices/ancillaries	E2402	Appropriate Clinical Signs, Symptoms and Conditions Term/Code not Available	A070801	Complete Loss of Power	28/07/2022	28/07/2022
							F26	No Health Consequences or Impact	A070801	Complete Loss of Power	28/07/2022	28/07/2022
					CT2750	Implantable cardiac electrical stimulators and associated devices	E2402	Appropriate Clinical Signs, Symptoms and Conditions Term/Code not Available	A070801	Complete Loss of Power	28/07/2022	28/07/2022
							F26	No Health Consequences or Impact	A070801	Complete Loss of Power	28/07/2022	28/07/2022
					CT2885	Physiological monitoring/measuring system transmitters/receivers/transceivers	E2402	Appropriate Clinical Signs, Symptoms and Conditions Term/Code not Available	A070801	Complete Loss of Power	28/07/2022	28/07/2022
							F26	No Health Consequences or Impact	A070801	Complete Loss of Power	28/07/2022	28/07/2022
12616	2023		47267	Single-chamber implantable pacemaker, rate-responsive	CT141	Electrodes/Leads and associated devices	E0602	Cardiac Arrest	A071204	No Pacing		
							F1205	Temporary Impairment	A071204	No Pacing		
					CT145	Cardiovascular devices	E0602	Cardiac Arrest	A071204	No Pacing		
							F1205	Temporary Impairment	A071204	No Pacing		
					CT1684	Cardiac pacemakers and associated devices	E0602	Cardiac Arrest	A071204	No Pacing		
							F1205	Temporary Impairment	A071204	No Pacing		
					CT1685	Cardiac pacemakers	E0602	Cardiac Arrest	A071204	No Pacing		
							F1205	Temporary Impairment	A071204	No Pacing		
					CT233	Surgical	E0602	Cardiac Arrest	A071204	No Pacing		
							F1205	Temporary Impairment	A071204	No Pacing		
					CT2406	Implantable	E0602	Cardiac Arrest	A071204	No Pacing		
							F1205	Temporary Impairment	A071204	No Pacing		
					CT2407	Active implantable	E0602	Cardiac Arrest	A071204	No Pacing		
							F1205	Temporary Impairment	A071204	No Pacing		
					CT257	Implantable cardiac pacemakers	E0602	Cardiac Arrest	A071204	No Pacing		
							F1205	Temporary Impairment	A071204	No Pacing		
					CT2727	Active implant devices/ancillaries	E0602	Cardiac Arrest	A071204	No Pacing		
							F1205	Temporary Impairment	A071204	No Pacing		
					CT2750	Implantable cardiac electrical stimulators and associated devices	E0602	Cardiac Arrest	A071204	No Pacing		
							F1205	Temporary Impairment	A071204	No Pacing		
					CT2866	Implantable cardiac electrical stimulators	E0602	Cardiac Arrest	A071204	No Pacing		
							F1205	Temporary Impairment	A071204	No Pacing		
					CT2985	Active	E0602	Cardiac Arrest	A071204	No Pacing		
							F1205	Temporary Impairment	A071204	No Pacing		
					CT3001	Implantable metal	E0602	Cardiac Arrest	A071204	No Pacing		
							F1205	Temporary Impairment	A071204	No Pacing		
					CT3002	Implantable synthetic polymer	E0602	Cardiac Arrest	A071204	No Pacing		
							F1205	Temporary Impairment	A071204	No Pacing		
					CT3006	Implantable inorganic material	E0602	Cardiac Arrest	A071204	No Pacing		
							F1205	Temporary Impairment	A071204	No Pacing		
					CT314	Electrically powered	E0602	Cardiac Arrest	A071204	No Pacing		
							F1205	Temporary Impairment	A071204	No Pacing		
					CT321	Long-term surgical invasive	E0602	Cardiac Arrest	A071204	No Pacing		
							F1205	Temporary Impairment	A071204	No Pacing		
					CT449	Generators and associated devices	E0602	Cardiac Arrest	A071204	No Pacing		
							F1205	Temporary Impairment	A071204	No Pacing		
					CT981	Single use	E0602	Cardiac Arrest	A071204	No Pacing		
							F1205	Temporary Impairment	A071204	No Pacing		
					CT983	Surgical invasive	E0602	Cardiac Arrest	A071204	No Pacing		
							F1205	Temporary Impairment	A071204	No Pacing		
12617	2023	All other reportable incidents	17882	Physiologic-monitoring defibrillation system	CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A071301	Failure to Deliver Shock/Stimulation		
							F26	No Health Consequences or Impact	A071301	Failure to Deliver Shock/Stimulation		
12618	2023	All other reportable incidents	17882	Physiologic-monitoring defibrillation system	CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A090402	Failure to Deliver Energy		
							F27	Problem Identified During Non-Clinical Procedure	A090402	Failure to Deliver Energy		
12619	2023	All other reportable incidents	17882	Physiologic-monitoring defibrillation system	CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A07	Electrical /Electronic Property Problem		
							F27	Problem Identified During Non-Clinical Procedure	A07	Electrical /Electronic Property Problem		
12620	2023	All other reportable incidents	35852	Single-chamber implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
									A1301	Failure to Read Input Signal		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
									A1301	Failure to Read Input Signal		
12621	2023	All other reportable incidents	47265	Dual-chamber implantable pacemaker, rate-responsive	CT141	Electrodes/Leads and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT145	Cardiovascular devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		

						F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery			
				CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery			
						F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery			
				CT233	Surgical	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery			
						F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery			
				CT2406	Implantable	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery			
						F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery			
				CT2407	Active implantable	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery			
						F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery			
				CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery			
						F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery			
				CT2727	Active implant devices/ancillaries	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery			
						F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery			
				CT2750	Implantable cardiac electrical stimulators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery			
						F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery			
				CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery			
						F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery			
				CT2985	Active	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery			
						F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery			
				CT3001	Implantable metal	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery			
						F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery			
				CT3002	Implantable synthetic polymer	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery			
						F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery			
				CT3006	Implantable inorganic material	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery			
						F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery			
				CT314	Electrically powered	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery			
						F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery			
				CT321	Long-term surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery			
						F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery			
				CT449	Generators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery			
						F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery			
				CT981	Single use	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery			
						F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery			
				CT983	Surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery			
						F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery			
12622	2023	All other reportable incidents	47263	Cardiac resynchronization therapy implantable pacemaker	CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A0908	Incorrect, Inadequate or Imprecise Result or Readings		
							F26	No Health Consequences or Impact	A0908	Incorrect, Inadequate or Imprecise Result or Readings		
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A0908	Incorrect, Inadequate or Imprecise Result or Readings		
							F26	No Health Consequences or Impact	A0908	Incorrect, Inadequate or Imprecise Result or Readings		
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A0908	Incorrect, Inadequate or Imprecise Result or Readings		
							F26	No Health Consequences or Impact	A0908	Incorrect, Inadequate or Imprecise Result or Readings		
12623	2023	All other reportable incidents	47263	Cardiac resynchronization therapy implantable pacemaker	CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	10/03/2023	10/03/2023
							F1901	Additional Surgery	A070504	Premature Discharge of Battery	10/03/2023	10/03/2023
					CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	10/03/2023	10/03/2023
							F1901	Additional Surgery	A070504	Premature Discharge of Battery	10/03/2023	10/03/2023
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	10/03/2023	10/03/2023
							F1901	Additional Surgery	A070504	Premature Discharge of Battery	10/03/2023	10/03/2023
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	10/03/2023	10/03/2023
							F1901	Additional Surgery	A070504	Premature Discharge of Battery	10/03/2023	10/03/2023
12624	2023	All other reportable incidents	47265	Dual-chamber implantable pacemaker, rate-responsive	CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem	13/03/2023	13/03/2023
									A0908	Incorrect, Inadequate or Imprecise Result or Readings	13/03/2023	13/03/2023
									A1304	Telemetry Discrepancy	13/03/2023	13/03/2023
							F08	Hospitalization or Prolonged Hospitalization	A0705	Battery Problem	13/03/2023	13/03/2023
									A0908	Incorrect, Inadequate or Imprecise Result or Readings	13/03/2023	13/03/2023
									A1304	Telemetry Discrepancy	13/03/2023	13/03/2023
							F1905	Device Revision or Replacement	A0705	Battery Problem	13/03/2023	13/03/2023
									A0908	Incorrect, Inadequate or Imprecise Result or Readings	13/03/2023	13/03/2023
									A1304	Telemetry Discrepancy	13/03/2023	13/03/2023
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem	13/03/2023	13/03/2023
									A0908	Incorrect, Inadequate or Imprecise Result or Readings	13/03/2023	13/03/2023
									A1304	Telemetry Discrepancy	13/03/2023	13/03/2023
							F08	Hospitalization or Prolonged Hospitalization	A0705	Battery Problem	13/03/2023	13/03/2023
									A0908	Incorrect, Inadequate or Imprecise Result or Readings	13/03/2023	13/03/2023
									A1304	Telemetry Discrepancy	13/03/2023	13/03/2023
							F1905	Device Revision or Replacement	A0705	Battery Problem	13/03/2023	13/03/2023
									A0908	Incorrect, Inadequate or Imprecise Result or Readings	13/03/2023	13/03/2023
									A1304	Telemetry Discrepancy	13/03/2023	13/03/2023
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem	13/03/2023	13/03/2023
									A0908	Incorrect, Inadequate or Imprecise Result or Readings	13/03/2023	13/03/2023
									A1304	Telemetry Discrepancy	13/03/2023	13/03/2023
							F08	Hospitalization or Prolonged Hospitalization	A0705	Battery Problem	13/03/2023	13/03/2023
									A0908	Incorrect, Inadequate or Imprecise Result or Readings	13/03/2023	13/03/2023
									A1304	Telemetry Discrepancy	13/03/2023	13/03/2023
							F1905	Device Revision or Replacement	A0705	Battery Problem	13/03/2023	13/03/2023
									A0908	Incorrect, Inadequate or Imprecise Result or Readings	13/03/2023	13/03/2023
									A1304	Telemetry Discrepancy	13/03/2023	13/03/2023

12625	2023		35852	Single-chamber implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A25	No Apparent Adverse Event		
							F26	No Health Consequences or Impact	A25	No Apparent Adverse Event		
12626	2023	All other reportable incidents	12913	Pacemaker, implantable	CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A070910	Undersensing		
									A1301	Failure to Read Input Signal		
							F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A070910	Undersensing		
									A1301	Failure to Read Input Signal		
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A070910	Undersensing		
									A1301	Failure to Read Input Signal		
							F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A070910	Undersensing		
									A1301	Failure to Read Input Signal		
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A070910	Undersensing		
									A1301	Failure to Read Input Signal		
							F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A070910	Undersensing		
									A1301	Failure to Read Input Signal		
12627	2023	All other reportable incidents	47263	Cardiac resynchronization therapy implantable pacemaker	CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A0908	Incorrect, Inadequate or Imprecise Result or Readings	24/08/2022	24/08/2022
							F08	Hospitalization or Prolonged Hospitalization	A0908	Incorrect, Inadequate or Imprecise Result or Readings	24/08/2022	24/08/2022
							F1905	Device Revision or Replacement	A0908	Incorrect, Inadequate or Imprecise Result or Readings	24/08/2022	24/08/2022
					CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A0908	Incorrect, Inadequate or Imprecise Result or Readings	24/08/2022	24/08/2022
							F08	Hospitalization or Prolonged Hospitalization	A0908	Incorrect, Inadequate or Imprecise Result or Readings	24/08/2022	24/08/2022
							F1905	Device Revision or Replacement	A0908	Incorrect, Inadequate or Imprecise Result or Readings	24/08/2022	24/08/2022
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A0908	Incorrect, Inadequate or Imprecise Result or Readings	24/08/2022	24/08/2022
							F08	Hospitalization or Prolonged Hospitalization	A0908	Incorrect, Inadequate or Imprecise Result or Readings	24/08/2022	24/08/2022
							F1905	Device Revision or Replacement	A0908	Incorrect, Inadequate or Imprecise Result or Readings	24/08/2022	24/08/2022
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A0908	Incorrect, Inadequate or Imprecise Result or Readings	24/08/2022	24/08/2022
							F08	Hospitalization or Prolonged Hospitalization	A0908	Incorrect, Inadequate or Imprecise Result or Readings	24/08/2022	24/08/2022
							F1905	Device Revision or Replacement	A0908	Incorrect, Inadequate or Imprecise Result or Readings	24/08/2022	24/08/2022
12628	2023	All other reportable incidents	46842	Cardiac pulse generator analyser cable, single-use	CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070102	High Capture Threshold		
									A070909	Oversensing		
									A072201	High Impedance		
							F26	No Health Consequences or Impact	A070102	High Capture Threshold		
									A070909	Oversensing		
									A072201	High Impedance		
12629	2023		47265	Dual-chamber implantable pacemaker, rate-responsive	CT141	Electrodes/Leads and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	24/02/2023	24/02/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	24/02/2023	24/02/2023
					CT145	Cardiovascular devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	24/02/2023	24/02/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	24/02/2023	24/02/2023
					CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	24/02/2023	24/02/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	24/02/2023	24/02/2023
					CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	24/02/2023	24/02/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	24/02/2023	24/02/2023
					CT233	Surgical	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	24/02/2023	24/02/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	24/02/2023	24/02/2023
					CT2406	Implantable	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	24/02/2023	24/02/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	24/02/2023	24/02/2023
					CT2407	Active implantable	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	24/02/2023	24/02/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	24/02/2023	24/02/2023
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	24/02/2023	24/02/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	24/02/2023	24/02/2023
					CT2727	Active implant devices/ancillaries	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	24/02/2023	24/02/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	24/02/2023	24/02/2023
					CT2750	Implantable cardiac electrical stimulators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	24/02/2023	24/02/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	24/02/2023	24/02/2023
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	24/02/2023	24/02/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	24/02/2023	24/02/2023
					CT2985	Active	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	24/02/2023	24/02/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	24/02/2023	24/02/2023
					CT3001	Implantable metal	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	24/02/2023	24/02/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	24/02/2023	24/02/2023
					CT3002	Implantable synthetic polymer	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	24/02/2023	24/02/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	24/02/2023	24/02/2023
					CT3006	Implantable inorganic material	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	24/02/2023	24/02/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	24/02/2023	24/02/2023
					CT314	Electrically powered	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	24/02/2023	24/02/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	24/02/2023	24/02/2023
					CT321	Long-term surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	24/02/2023	24/02/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	24/02/2023	24/02/2023
					CT449	Generators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	24/02/2023	24/02/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	24/02/2023	24/02/2023
					CT981	Single use	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	24/02/2023	24/02/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	24/02/2023	24/02/2023
					CT983	Surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	24/02/2023	24/02/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	24/02/2023	24/02/2023
12630	2023	All other reportable incidents	35852	Single-chamber implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E0119	Loss of Consciousness	A070909	Oversensing		
									A071201	Failure to Convert Rhythm		
									A072201	High Impedance		
							E011903	Syncope/Fainting	A070909	Oversensing		
									A071201	Failure to Convert Rhythm		
									A072201	High Impedance		
							F08	Hospitalization or Prolonged Hospitalization	A070909	Oversensing		

									A071201	Failure to Convert Rhythm		
									A072201	High Impedance		
						F2203	Imaging Required	A070909	Oversensing			
								A071201	Failure to Convert Rhythm			
								A072201	High Impedance			
						F2306	Resuscitation	A070909	Oversensing			
								A071201	Failure to Convert Rhythm			
								A072201	High Impedance			
12631	2023	All other reportable incidents	47265	Dual-chamber implantable pacemaker, rate-responsive	CT141	Electrodes/Leads and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
					CT145	Cardiovascular devices	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
					CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
					CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
					CT233	Surgical	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
					CT2406	Implantable	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
					CT2407	Active implantable	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
					CT2727	Active implant devices/ancillaries	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
					CT2750	Implantable cardiac electrical stimulators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
					CT2985	Active	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
					CT3001	Implantable metal	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
					CT3002	Implantable synthetic polymer	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
					CT3006	Implantable inorganic material	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
					CT314	Electrically powered	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
					CT321	Long-term surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
					CT449	Generators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
					CT981	Single use	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
					CT983	Surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		

12632	2023		47265	Dual-chamber implantable pacemaker, rate-responsive	CT141	Electrodes/Leads and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	06/03/2023	06/03/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	06/03/2023	06/03/2023
					CT145	Cardiovascular devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	06/03/2023	06/03/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	06/03/2023	06/03/2023
					CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	06/03/2023	06/03/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	06/03/2023	06/03/2023
					CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	06/03/2023	06/03/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	06/03/2023	06/03/2023
					CT233	Surgical	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	06/03/2023	06/03/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	06/03/2023	06/03/2023
					CT2406	Implantable	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	06/03/2023	06/03/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	06/03/2023	06/03/2023
					CT2407	Active implantable	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	06/03/2023	06/03/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	06/03/2023	06/03/2023
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	06/03/2023	06/03/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	06/03/2023	06/03/2023
					CT2727	Active implant devices/ancillaries	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	06/03/2023	06/03/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	06/03/2023	06/03/2023
					CT2750	Implantable cardiac electrical stimulators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	06/03/2023	06/03/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	06/03/2023	06/03/2023
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	06/03/2023	06/03/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	06/03/2023	06/03/2023
					CT2985	Active	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	06/03/2023	06/03/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	06/03/2023	06/03/2023
					CT3001	Implantable metal	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	06/03/2023	06/03/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	06/03/2023	06/03/2023
					CT3002	Implantable synthetic polymer	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	06/03/2023	06/03/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	06/03/2023	06/03/2023
					CT3006	Implantable inorganic material	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	06/03/2023	06/03/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	06/03/2023	06/03/2023
					CT314	Electrically powered	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	06/03/2023	06/03/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	06/03/2023	06/03/2023
					CT321	Long-term surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	06/03/2023	06/03/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	06/03/2023	06/03/2023
					CT449	Generators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	06/03/2023	06/03/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	06/03/2023	06/03/2023
					CT981	Single use	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	06/03/2023	06/03/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	06/03/2023	06/03/2023
					CT983	Surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	06/03/2023	06/03/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	06/03/2023	06/03/2023
12633	2023		47265	Dual-chamber implantable pacemaker, rate-responsive	CT141	Electrodes/Leads and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	03/02/2023	03/02/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	03/02/2023	03/02/2023
					CT145	Cardiovascular devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	03/02/2023	03/02/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	03/02/2023	03/02/2023
					CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	03/02/2023	03/02/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	03/02/2023	03/02/2023
					CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	03/02/2023	03/02/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	03/02/2023	03/02/2023
					CT233	Surgical	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	03/02/2023	03/02/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	03/02/2023	03/02/2023
					CT2406	Implantable	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	03/02/2023	03/02/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	03/02/2023	03/02/2023
					CT2407	Active implantable	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	03/02/2023	03/02/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	03/02/2023	03/02/2023
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	03/02/2023	03/02/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	03/02/2023	03/02/2023
					CT2727	Active implant devices/ancillaries	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	03/02/2023	03/02/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	03/02/2023	03/02/2023
					CT2750	Implantable cardiac electrical stimulators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	03/02/2023	03/02/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	03/02/2023	03/02/2023
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	03/02/2023	03/02/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	03/02/2023	03/02/2023
					CT2985	Active	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	03/02/2023	03/02/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	03/02/2023	03/02/2023
					CT3001	Implantable metal	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	03/02/2023	03/02/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	03/02/2023	03/02/2023
					CT3002	Implantable synthetic polymer	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	03/02/2023	03/02/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	03/02/2023	03/02/2023
					CT3006	Implantable inorganic material	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	03/02/2023	03/02/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	03/02/2023	03/02/2023
					CT314	Electrically powered	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	03/02/2023	03/02/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	03/02/2023	03/02/2023
					CT321	Long-term surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	03/02/2023	03/02/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	03/02/2023	03/02/2023
					CT449	Generators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	03/02/2023	03/02/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	03/02/2023	03/02/2023
					CT981	Single use	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	03/02/2023	03/02/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	03/02/2023	03/02/2023
					CT983	Surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	03/02/2023	03/02/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	03/02/2023	03/02/2023
12634	2023		47265	Dual-chamber implantable pacemaker, rate-responsive	CT141	Electrodes/Leads and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	27/01/2023	27/01/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	27/01/2023	27/01/2023
					CT145	Cardiovascular devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	27/01/2023	27/01/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	27/01/2023	27/01/2023
					CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	27/01/2023	27/01/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	27/01/2023	27/01/2023
					CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	27/01/2023	27/01/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	27/01/2023	27/01/2023

					CT233	Surgical	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	27/01/2023	27/01/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	27/01/2023	27/01/2023
					CT2406	Implantable	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	27/01/2023	27/01/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	27/01/2023	27/01/2023
					CT2407	Active implantable	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	27/01/2023	27/01/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	27/01/2023	27/01/2023
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	27/01/2023	27/01/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	27/01/2023	27/01/2023
					CT2727	Active implant devices/ancillaries	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	27/01/2023	27/01/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	27/01/2023	27/01/2023
					CT2750	Implantable cardiac electrical stimulators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	27/01/2023	27/01/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	27/01/2023	27/01/2023
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	27/01/2023	27/01/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	27/01/2023	27/01/2023
					CT2985	Active	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	27/01/2023	27/01/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	27/01/2023	27/01/2023
					CT3001	Implantable metal	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	27/01/2023	27/01/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	27/01/2023	27/01/2023
					CT3002	Implantable synthetic polymer	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	27/01/2023	27/01/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	27/01/2023	27/01/2023
					CT3006	Implantable inorganic material	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	27/01/2023	27/01/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	27/01/2023	27/01/2023
					CT314	Electrically powered	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	27/01/2023	27/01/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	27/01/2023	27/01/2023
					CT321	Long-term surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	27/01/2023	27/01/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	27/01/2023	27/01/2023
					CT449	Generators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	27/01/2023	27/01/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	27/01/2023	27/01/2023
					CT981	Single use	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	27/01/2023	27/01/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	27/01/2023	27/01/2023
					CT983	Surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	27/01/2023	27/01/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	27/01/2023	27/01/2023
12635	2023	All other reportable incidents	47265	Dual-chamber implantable pacemaker, rate-responsive	CT141	Electrodes/Leads and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	07/03/2023	07/03/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	07/03/2023	07/03/2023
					CT145	Cardiovascular devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	07/03/2023	07/03/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	07/03/2023	07/03/2023
					CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	07/03/2023	07/03/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	07/03/2023	07/03/2023
					CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	07/03/2023	07/03/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	07/03/2023	07/03/2023
					CT233	Surgical	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	07/03/2023	07/03/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	07/03/2023	07/03/2023
					CT2406	Implantable	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	07/03/2023	07/03/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	07/03/2023	07/03/2023
					CT2407	Active implantable	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	07/03/2023	07/03/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	07/03/2023	07/03/2023
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	07/03/2023	07/03/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	07/03/2023	07/03/2023
					CT2727	Active implant devices/ancillaries	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	07/03/2023	07/03/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	07/03/2023	07/03/2023
					CT2750	Implantable cardiac electrical stimulators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	07/03/2023	07/03/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	07/03/2023	07/03/2023
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	07/03/2023	07/03/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	07/03/2023	07/03/2023
					CT2985	Active	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	07/03/2023	07/03/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	07/03/2023	07/03/2023
					CT3001	Implantable metal	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	07/03/2023	07/03/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	07/03/2023	07/03/2023
					CT3002	Implantable synthetic polymer	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	07/03/2023	07/03/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	07/03/2023	07/03/2023
					CT3006	Implantable inorganic material	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	07/03/2023	07/03/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	07/03/2023	07/03/2023
					CT314	Electrically powered	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	07/03/2023	07/03/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	07/03/2023	07/03/2023
					CT321	Long-term surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	07/03/2023	07/03/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	07/03/2023	07/03/2023
					CT449	Generators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	07/03/2023	07/03/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	07/03/2023	07/03/2023
					CT981	Single use	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	07/03/2023	07/03/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	07/03/2023	07/03/2023
					CT983	Surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	07/03/2023	07/03/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	07/03/2023	07/03/2023
12636	2023		47265	Dual-chamber implantable pacemaker, rate-responsive	CT141	Electrodes/Leads and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	29/01/2023	29/01/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	29/01/2023	29/01/2023
					CT145	Cardiovascular devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	29/01/2023	29/01/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	29/01/2023	29/01/2023
					CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	29/01/2023	29/01/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	29/01/2023	29/01/2023
					CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	29/01/2023	29/01/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	29/01/2023	29/01/2023
					CT233	Surgical	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	29/01/2023	29/01/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	29/01/2023	29/01/2023
					CT2406	Implantable	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	29/01/2023	29/01/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	29/01/2023	29/01/2023
					CT2407	Active implantable	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	29/01/2023	29/01/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	29/01/2023	29/01/2023
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	29/01/2023	29/01/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	29/01/2023	29/01/2023

					CT2727	Active implant devices/ancillaries	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	29/01/2023	29/01/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	29/01/2023	29/01/2023
					CT2750	Implantable cardiac electrical stimulators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	29/01/2023	29/01/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	29/01/2023	29/01/2023
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	29/01/2023	29/01/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	29/01/2023	29/01/2023
					CT2985	Active	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	29/01/2023	29/01/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	29/01/2023	29/01/2023
					CT3001	Implantable metal	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	29/01/2023	29/01/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	29/01/2023	29/01/2023
					CT3002	Implantable synthetic polymer	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	29/01/2023	29/01/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	29/01/2023	29/01/2023
					CT3006	Implantable inorganic material	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	29/01/2023	29/01/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	29/01/2023	29/01/2023
					CT314	Electrically powered	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	29/01/2023	29/01/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	29/01/2023	29/01/2023
					CT321	Long-term surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	29/01/2023	29/01/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	29/01/2023	29/01/2023
					CT449	Generators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	29/01/2023	29/01/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	29/01/2023	29/01/2023
					CT981	Single use	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	29/01/2023	29/01/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	29/01/2023	29/01/2023
					CT983	Surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	29/01/2023	29/01/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	29/01/2023	29/01/2023
12637	2023	All other reportable incidents	17882	Physiologic-monitoring defibrillation system	CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A0801	Failure to Calibrate		
							F27	Problem Identified During Non-Clinical Procedure	A0801	Failure to Calibrate		
12638	2023	All other reportable incidents	47265	Dual-chamber implantable pacemaker, rate-responsive	CT141	Electrodes/Leads and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT145	Cardiovascular devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT233	Surgical	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT2406	Implantable	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT2407	Active implantable	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT2727	Active implant devices/ancillaries	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT2750	Implantable cardiac electrical stimulators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT2985	Active	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT3001	Implantable metal	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT3002	Implantable synthetic polymer	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT3006	Implantable inorganic material	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT314	Electrically powered	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT321	Long-term surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT449	Generators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT981	Single use	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT983	Surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
12639	2023	All other reportable incidents	47265	Dual-chamber implantable pacemaker, rate-responsive	CT141	Electrodes/Leads and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	17/03/2023	17/03/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	17/03/2023	17/03/2023
					CT145	Cardiovascular devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	17/03/2023	17/03/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	17/03/2023	17/03/2023
					CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	17/03/2023	17/03/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	17/03/2023	17/03/2023
					CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	17/03/2023	17/03/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	17/03/2023	17/03/2023
					CT233	Surgical	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	17/03/2023	17/03/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	17/03/2023	17/03/2023
					CT2406	Implantable	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	17/03/2023	17/03/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	17/03/2023	17/03/2023
					CT2407	Active implantable	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	17/03/2023	17/03/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	17/03/2023	17/03/2023
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	17/03/2023	17/03/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	17/03/2023	17/03/2023
					CT2727	Active implant devices/ancillaries	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	17/03/2023	17/03/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	17/03/2023	17/03/2023
					CT2750	Implantable cardiac electrical stimulators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	17/03/2023	17/03/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	17/03/2023	17/03/2023
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	17/03/2023	17/03/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	17/03/2023	17/03/2023

					CT2985	Active	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	17/03/2023	17/03/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	17/03/2023	17/03/2023
					CT3001	Implantable metal	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	17/03/2023	17/03/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	17/03/2023	17/03/2023
					CT3002	Implantable synthetic polymer	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	17/03/2023	17/03/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	17/03/2023	17/03/2023
					CT3006	Implantable inorganic material	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	17/03/2023	17/03/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	17/03/2023	17/03/2023
					CT314	Electrically powered	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	17/03/2023	17/03/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	17/03/2023	17/03/2023
					CT321	Long-term surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	17/03/2023	17/03/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	17/03/2023	17/03/2023
					CT449	Generators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	17/03/2023	17/03/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	17/03/2023	17/03/2023
					CT981	Single use	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	17/03/2023	17/03/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	17/03/2023	17/03/2023
					CT983	Surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	17/03/2023	17/03/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	17/03/2023	17/03/2023
12640	2023		47265	Dual-chamber implantable pacemaker, rate-responsive	CT141	Electrodes/Leads and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	03/03/2023	03/03/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	03/03/2023	03/03/2023
					CT145	Cardiovascular devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	03/03/2023	03/03/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	03/03/2023	03/03/2023
					CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	03/03/2023	03/03/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	03/03/2023	03/03/2023
					CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	03/03/2023	03/03/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	03/03/2023	03/03/2023
					CT233	Surgical	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	03/03/2023	03/03/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	03/03/2023	03/03/2023
					CT2406	Implantable	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	03/03/2023	03/03/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	03/03/2023	03/03/2023
					CT2407	Active implantable	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	03/03/2023	03/03/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	03/03/2023	03/03/2023
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	03/03/2023	03/03/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	03/03/2023	03/03/2023
					CT2727	Active implant devices/ancillaries	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	03/03/2023	03/03/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	03/03/2023	03/03/2023
					CT2750	Implantable cardiac electrical stimulators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	03/03/2023	03/03/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	03/03/2023	03/03/2023
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	03/03/2023	03/03/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	03/03/2023	03/03/2023
					CT2985	Active	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	03/03/2023	03/03/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	03/03/2023	03/03/2023
					CT3001	Implantable metal	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	03/03/2023	03/03/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	03/03/2023	03/03/2023
					CT3002	Implantable synthetic polymer	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	03/03/2023	03/03/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	03/03/2023	03/03/2023
					CT3006	Implantable inorganic material	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	03/03/2023	03/03/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	03/03/2023	03/03/2023
					CT314	Electrically powered	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	03/03/2023	03/03/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	03/03/2023	03/03/2023
					CT321	Long-term surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	03/03/2023	03/03/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	03/03/2023	03/03/2023
					CT449	Generators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	03/03/2023	03/03/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	03/03/2023	03/03/2023
					CT981	Single use	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	03/03/2023	03/03/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	03/03/2023	03/03/2023
					CT983	Surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	03/03/2023	03/03/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	03/03/2023	03/03/2023
12641	2023	All other reportable incidents	17882	Physiologic-monitoring defibrillation system	CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A0719	Unexpected Shutdown		
							F27	Problem Identified During Non-Clinical Procedure	A0719	Unexpected Shutdown		
12642	2023		47265	Dual-chamber implantable pacemaker, rate-responsive	CT141	Electrodes/Leads and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem		
							F1901	Additional Surgery	A24	No Apparent Device Problem		
					CT145	Cardiovascular devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem		
							F1901	Additional Surgery	A24	No Apparent Device Problem		
					CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem		
							F1901	Additional Surgery	A24	No Apparent Device Problem		
					CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem		
							F1901	Additional Surgery	A24	No Apparent Device Problem		
					CT233	Surgical	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem		
							F1901	Additional Surgery	A24	No Apparent Device Problem		
					CT2406	Implantable	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem		
							F1901	Additional Surgery	A24	No Apparent Device Problem		
					CT2407	Active implantable	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem		
							F1901	Additional Surgery	A24	No Apparent Device Problem		
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem		
							F1901	Additional Surgery	A24	No Apparent Device Problem		
					CT2727	Active implant devices/ancillaries	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem		
							F1901	Additional Surgery	A24	No Apparent Device Problem		
					CT2750	Implantable cardiac electrical stimulators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem		
							F1901	Additional Surgery	A24	No Apparent Device Problem		
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem		
							F1901	Additional Surgery	A24	No Apparent Device Problem		
					CT2985	Active	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem		
							F1901	Additional Surgery	A24	No Apparent Device Problem		
					CT3001	Implantable metal	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem		
							F1901	Additional Surgery	A24	No Apparent Device Problem		
					CT3002	Implantable synthetic polymer	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem		
							F1901	Additional Surgery	A24	No Apparent Device Problem		

					CT3006	Implantable inorganic material	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem		
							F1901	Additional Surgery	A24	No Apparent Device Problem		
					CT314	Electrically powered	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem		
							F1901	Additional Surgery	A24	No Apparent Device Problem		
					CT321	Long-term surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem		
							F1901	Additional Surgery	A24	No Apparent Device Problem		
					CT449	Generators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem		
							F1901	Additional Surgery	A24	No Apparent Device Problem		
					CT981	Single use	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem		
							F1901	Additional Surgery	A24	No Apparent Device Problem		
					CT983	Surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem		
							F1901	Additional Surgery	A24	No Apparent Device Problem		
12643	2023	All other reportable incidents	37265	Dual-chamber implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A070909	Oversensing		
							F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A070909	Oversensing		
12644	2023	All other reportable incidents	47263	Cardiac resynchronization therapy implantable pacemaker	CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A0908	Incorrect, Inadequate or Imprecise Result or Readings	23/11/2022	23/11/2022
							F08	Hospitalization or Prolonged Hospitalization	A0908	Incorrect, Inadequate or Imprecise Result or Readings	23/11/2022	23/11/2022
							F1905	Device Revision or Replacement	A0908	Incorrect, Inadequate or Imprecise Result or Readings	23/11/2022	23/11/2022
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A0908	Incorrect, Inadequate or Imprecise Result or Readings	23/11/2022	23/11/2022
							F08	Hospitalization or Prolonged Hospitalization	A0908	Incorrect, Inadequate or Imprecise Result or Readings	23/11/2022	23/11/2022
							F1905	Device Revision or Replacement	A0908	Incorrect, Inadequate or Imprecise Result or Readings	23/11/2022	23/11/2022
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A0908	Incorrect, Inadequate or Imprecise Result or Readings	23/11/2022	23/11/2022
							F08	Hospitalization or Prolonged Hospitalization	A0908	Incorrect, Inadequate or Imprecise Result or Readings	23/11/2022	23/11/2022
							F1905	Device Revision or Replacement	A0908	Incorrect, Inadequate or Imprecise Result or Readings	23/11/2022	23/11/2022
12645	2023	All other reportable incidents	37265	Dual-chamber implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
12646	2023	All other reportable incidents	17882	Physiologic-monitoring defibrillation system	CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A090102	Inaudible or Unclear Audible Prompt/Feedback		
							F27	Problem Identified During Non-Clinical Procedure	A090102	Inaudible or Unclear Audible Prompt/Feedback		
12647	2023		47265	Dual-chamber implantable pacemaker, rate-responsive	CT141	Electrodes/Leads and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem		
							F1901	Additional Surgery	A24	No Apparent Device Problem		
					CT145	Cardiovascular devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem		
							F1901	Additional Surgery	A24	No Apparent Device Problem		
					CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem		
							F1901	Additional Surgery	A24	No Apparent Device Problem		
					CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem		
							F1901	Additional Surgery	A24	No Apparent Device Problem		
					CT233	Surgical	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem		
							F1901	Additional Surgery	A24	No Apparent Device Problem		
					CT2406	Implantable	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem		
							F1901	Additional Surgery	A24	No Apparent Device Problem		
					CT2407	Active implantable	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem		
							F1901	Additional Surgery	A24	No Apparent Device Problem		
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem		
							F1901	Additional Surgery	A24	No Apparent Device Problem		
					CT2727	Active implant devices/ancillaries	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem		
							F1901	Additional Surgery	A24	No Apparent Device Problem		
					CT2750	Implantable cardiac electrical stimulators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem		
							F1901	Additional Surgery	A24	No Apparent Device Problem		
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem		
							F1901	Additional Surgery	A24	No Apparent Device Problem		
					CT2985	Active	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem		
							F1901	Additional Surgery	A24	No Apparent Device Problem		
					CT3001	Implantable metal	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem		
							F1901	Additional Surgery	A24	No Apparent Device Problem		
					CT3002	Implantable synthetic polymer	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem		
							F1901	Additional Surgery	A24	No Apparent Device Problem		
					CT3006	Implantable inorganic material	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem		
							F1901	Additional Surgery	A24	No Apparent Device Problem		
					CT314	Electrically powered	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem		
							F1901	Additional Surgery	A24	No Apparent Device Problem		
					CT321	Long-term surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem		
							F1901	Additional Surgery	A24	No Apparent Device Problem		
					CT449	Generators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem		
							F1901	Additional Surgery	A24	No Apparent Device Problem		
					CT981	Single use	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem		
							F1901	Additional Surgery	A24	No Apparent Device Problem		
					CT983	Surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem		
							F1901	Additional Surgery	A24	No Apparent Device Problem		
12648	2023	All other reportable incidents	17882	Physiologic-monitoring defibrillation system	CT1684	Cardiac pacemakers and associated devices	E2401	Insufficient Information	A071204	No Pacing		
							F24	Insufficient Health Impact Information	A071204	No Pacing		
12649	2023	All other reportable incidents	36241	Pacemaker/defibrillator lead	CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A051201	Device Dislodged or Dislocated		
							F2203	Imaging Required	A051201	Device Dislodged or Dislocated		
12650	2023	All other reportable incidents	47263	Cardiac resynchronization therapy implantable pacemaker	CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A0908	Incorrect, Inadequate or Imprecise Result or Readings	08/03/2023	08/03/2023
							F08	Hospitalization or Prolonged Hospitalization	A0908	Incorrect, Inadequate or Imprecise Result or Readings	08/03/2023	08/03/2023
							F1905	Device Revision or Replacement	A0908	Incorrect, Inadequate or Imprecise Result or Readings	08/03/2023	08/03/2023
					CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A0908	Incorrect, Inadequate or Imprecise Result or Readings	08/03/2023	08/03/2023

							F08	Hospitalization or Prolonged Hospitalization	A0908	Incorrect, Inadequate or Imprecise Result or Readings	08/03/2023	08/03/2023
							F1905	Device Revision or Replacement	A0908	Incorrect, Inadequate or Imprecise Result or Readings	08/03/2023	08/03/2023
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A0908	Incorrect, Inadequate or Imprecise Result or Readings	08/03/2023	08/03/2023
							F08	Hospitalization or Prolonged Hospitalization	A0908	Incorrect, Inadequate or Imprecise Result or Readings	08/03/2023	08/03/2023
							F1905	Device Revision or Replacement	A0908	Incorrect, Inadequate or Imprecise Result or Readings	08/03/2023	08/03/2023
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A0908	Incorrect, Inadequate or Imprecise Result or Readings	08/03/2023	08/03/2023
							F08	Hospitalization or Prolonged Hospitalization	A0908	Incorrect, Inadequate or Imprecise Result or Readings	08/03/2023	08/03/2023
							F1905	Device Revision or Replacement	A0908	Incorrect, Inadequate or Imprecise Result or Readings	08/03/2023	08/03/2023
12651	2023	All other reportable incidents	47265	Dual-chamber implantable pacemaker, rate-responsive	CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
12652	2023		37265	Dual-chamber implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E0112	Dizziness	A070909	Oversensing	13/03/2023	13/03/2023
									A071204	No Pacing	13/03/2023	13/03/2023
							E0601	Arrhythmia	A070909	Oversensing	13/03/2023	13/03/2023
									A071204	No Pacing	13/03/2023	13/03/2023
							F1901	Additional Surgery	A070909	Oversensing	13/03/2023	13/03/2023
									A071204	No Pacing	13/03/2023	13/03/2023
12653	2023		47265	Dual-chamber implantable pacemaker, rate-responsive	CT141	Electrodes/Leads and associated devices	E0601	Arrhythmia	A071204	No Pacing	07/03/2023	07/03/2023
							F12	Serious Injury/ Illness/ Impairment	A071204	No Pacing	07/03/2023	07/03/2023
					CT145	Cardiovascular devices	E0601	Arrhythmia	A071204	No Pacing	07/03/2023	07/03/2023
							F12	Serious Injury/ Illness/ Impairment	A071204	No Pacing	07/03/2023	07/03/2023
					CT1684	Cardiac pacemakers and associated devices	E0601	Arrhythmia	A071204	No Pacing	07/03/2023	07/03/2023
							F12	Serious Injury/ Illness/ Impairment	A071204	No Pacing	07/03/2023	07/03/2023
					CT1685	Cardiac pacemakers	E0601	Arrhythmia	A071204	No Pacing	07/03/2023	07/03/2023
							F12	Serious Injury/ Illness/ Impairment	A071204	No Pacing	07/03/2023	07/03/2023
					CT233	Surgical	E0601	Arrhythmia	A071204	No Pacing	07/03/2023	07/03/2023
							F12	Serious Injury/ Illness/ Impairment	A071204	No Pacing	07/03/2023	07/03/2023
					CT2406	Implantable	E0601	Arrhythmia	A071204	No Pacing	07/03/2023	07/03/2023
							F12	Serious Injury/ Illness/ Impairment	A071204	No Pacing	07/03/2023	07/03/2023
					CT2407	Active implantable	E0601	Arrhythmia	A071204	No Pacing	07/03/2023	07/03/2023
							F12	Serious Injury/ Illness/ Impairment	A071204	No Pacing	07/03/2023	07/03/2023
					CT257	Implantable cardiac pacemakers	E0601	Arrhythmia	A071204	No Pacing	07/03/2023	07/03/2023
							F12	Serious Injury/ Illness/ Impairment	A071204	No Pacing	07/03/2023	07/03/2023
					CT2727	Active implant devices/ancillaries	E0601	Arrhythmia	A071204	No Pacing	07/03/2023	07/03/2023
							F12	Serious Injury/ Illness/ Impairment	A071204	No Pacing	07/03/2023	07/03/2023
					CT2750	Implantable cardiac electrical stimulators and associated devices	E0601	Arrhythmia	A071204	No Pacing	07/03/2023	07/03/2023
							F12	Serious Injury/ Illness/ Impairment	A071204	No Pacing	07/03/2023	07/03/2023
					CT2866	Implantable cardiac electrical stimulators	E0601	Arrhythmia	A071204	No Pacing	07/03/2023	07/03/2023
							F12	Serious Injury/ Illness/ Impairment	A071204	No Pacing	07/03/2023	07/03/2023
					CT2985	Active	E0601	Arrhythmia	A071204	No Pacing	07/03/2023	07/03/2023
							F12	Serious Injury/ Illness/ Impairment	A071204	No Pacing	07/03/2023	07/03/2023
					CT3001	Implantable metal	E0601	Arrhythmia	A071204	No Pacing	07/03/2023	07/03/2023
							F12	Serious Injury/ Illness/ Impairment	A071204	No Pacing	07/03/2023	07/03/2023
					CT3002	Implantable synthetic polymer	E0601	Arrhythmia	A071204	No Pacing	07/03/2023	07/03/2023
							F12	Serious Injury/ Illness/ Impairment	A071204	No Pacing	07/03/2023	07/03/2023
					CT3006	Implantable inorganic material	E0601	Arrhythmia	A071204	No Pacing	07/03/2023	07/03/2023
							F12	Serious Injury/ Illness/ Impairment	A071204	No Pacing	07/03/2023	07/03/2023
					CT314	Electrically powered	E0601	Arrhythmia	A071204	No Pacing	07/03/2023	07/03/2023
							F12	Serious Injury/ Illness/ Impairment	A071204	No Pacing	07/03/2023	07/03/2023
					CT321	Long-term surgical invasive	E0601	Arrhythmia	A071204	No Pacing	07/03/2023	07/03/2023
							F12	Serious Injury/ Illness/ Impairment	A071204	No Pacing	07/03/2023	07/03/2023
					CT449	Generators and associated devices	E0601	Arrhythmia	A071204	No Pacing	07/03/2023	07/03/2023
							F12	Serious Injury/ Illness/ Impairment	A071204	No Pacing	07/03/2023	07/03/2023
					CT981	Single use	E0601	Arrhythmia	A071204	No Pacing	07/03/2023	07/03/2023
							F12	Serious Injury/ Illness/ Impairment	A071204	No Pacing	07/03/2023	07/03/2023
					CT983	Surgical invasive	E0601	Arrhythmia	A071204	No Pacing	07/03/2023	07/03/2023
							F12	Serious Injury/ Illness/ Impairment	A071204	No Pacing	07/03/2023	07/03/2023
12654	2023		47265	Dual-chamber implantable pacemaker, rate-responsive	CT141	Electrodes/Leads and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	23/03/2023	23/03/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	23/03/2023	23/03/2023
					CT145	Cardiovascular devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	23/03/2023	23/03/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	23/03/2023	23/03/2023
					CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	23/03/2023	23/03/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	23/03/2023	23/03/2023
					CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	23/03/2023	23/03/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	23/03/2023	23/03/2023
					CT233	Surgical	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	23/03/2023	23/03/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	23/03/2023	23/03/2023
					CT2406	Implantable	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	23/03/2023	23/03/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	23/03/2023	23/03/2023
					CT2407	Active implantable	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	23/03/2023	23/03/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	23/03/2023	23/03/2023
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	23/03/2023	23/03/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	23/03/2023	23/03/2023
					CT2727	Active implant devices/ancillaries	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	23/03/2023	23/03/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	23/03/2023	23/03/2023
					CT2750	Implantable cardiac electrical stimulators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	23/03/2023	23/03/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	23/03/2023	23/03/2023
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	23/03/2023	23/03/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	23/03/2023	23/03/2023

					CT2985	Active	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	23/03/2023	23/03/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	23/03/2023	23/03/2023
					CT3001	Implantable metal	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	23/03/2023	23/03/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	23/03/2023	23/03/2023
					CT3002	Implantable synthetic polymer	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	23/03/2023	23/03/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	23/03/2023	23/03/2023
					CT3006	Implantable inorganic material	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	23/03/2023	23/03/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	23/03/2023	23/03/2023
					CT314	Electrically powered	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	23/03/2023	23/03/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	23/03/2023	23/03/2023
					CT321	Long-term surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	23/03/2023	23/03/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	23/03/2023	23/03/2023
					CT449	Generators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	23/03/2023	23/03/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	23/03/2023	23/03/2023
					CT981	Single use	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	23/03/2023	23/03/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	23/03/2023	23/03/2023
					CT983	Surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	23/03/2023	23/03/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	23/03/2023	23/03/2023
12655	2023		47265	Dual-chamber implantable pacemaker, rate-responsive	CT141	Electrodes/Leads and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	12/02/2023	12/02/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	12/02/2023	12/02/2023
					CT145	Cardiovascular devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	12/02/2023	12/02/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	12/02/2023	12/02/2023
					CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	12/02/2023	12/02/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	12/02/2023	12/02/2023
					CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	12/02/2023	12/02/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	12/02/2023	12/02/2023
					CT233	Surgical	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	12/02/2023	12/02/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	12/02/2023	12/02/2023
					CT2406	Implantable	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	12/02/2023	12/02/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	12/02/2023	12/02/2023
					CT2407	Active implantable	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	12/02/2023	12/02/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	12/02/2023	12/02/2023
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	12/02/2023	12/02/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	12/02/2023	12/02/2023
					CT2727	Active implant devices/ancillaries	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	12/02/2023	12/02/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	12/02/2023	12/02/2023
					CT2750	Implantable cardiac electrical stimulators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	12/02/2023	12/02/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	12/02/2023	12/02/2023
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	12/02/2023	12/02/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	12/02/2023	12/02/2023
					CT2985	Active	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	12/02/2023	12/02/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	12/02/2023	12/02/2023
					CT3001	Implantable metal	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	12/02/2023	12/02/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	12/02/2023	12/02/2023
					CT3002	Implantable synthetic polymer	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	12/02/2023	12/02/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	12/02/2023	12/02/2023
					CT3006	Implantable inorganic material	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	12/02/2023	12/02/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	12/02/2023	12/02/2023
					CT314	Electrically powered	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	12/02/2023	12/02/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	12/02/2023	12/02/2023
					CT321	Long-term surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	12/02/2023	12/02/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	12/02/2023	12/02/2023
					CT449	Generators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	12/02/2023	12/02/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	12/02/2023	12/02/2023
					CT981	Single use	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	12/02/2023	12/02/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	12/02/2023	12/02/2023
					CT983	Surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	12/02/2023	12/02/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	12/02/2023	12/02/2023
12656	2023		47265	Dual-chamber implantable pacemaker, rate-responsive	CT141	Electrodes/Leads and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	19/02/2023	19/02/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	19/02/2023	19/02/2023
					CT145	Cardiovascular devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	19/02/2023	19/02/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	19/02/2023	19/02/2023
					CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	19/02/2023	19/02/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	19/02/2023	19/02/2023
					CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	19/02/2023	19/02/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	19/02/2023	19/02/2023
					CT233	Surgical	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	19/02/2023	19/02/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	19/02/2023	19/02/2023
					CT2406	Implantable	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	19/02/2023	19/02/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	19/02/2023	19/02/2023
					CT2407	Active implantable	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	19/02/2023	19/02/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	19/02/2023	19/02/2023
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	19/02/2023	19/02/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	19/02/2023	19/02/2023
					CT2727	Active implant devices/ancillaries	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	19/02/2023	19/02/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	19/02/2023	19/02/2023
					CT2750	Implantable cardiac electrical stimulators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	19/02/2023	19/02/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	19/02/2023	19/02/2023
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	19/02/2023	19/02/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	19/02/2023	19/02/2023
					CT2985	Active	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	19/02/2023	19/02/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	19/02/2023	19/02/2023
					CT3001	Implantable metal	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	19/02/2023	19/02/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	19/02/2023	19/02/2023
					CT3002	Implantable synthetic polymer	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	19/02/2023	19/02/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	19/02/2023	19/02/2023
					CT3006	Implantable inorganic material	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	19/02/2023	19/02/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	19/02/2023	19/02/2023

					CT314	Electrically powered	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	19/02/2023	19/02/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	19/02/2023	19/02/2023
					CT321	Long-term surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	19/02/2023	19/02/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	19/02/2023	19/02/2023
					CT449	Generators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	19/02/2023	19/02/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	19/02/2023	19/02/2023
					CT981	Single use	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	19/02/2023	19/02/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	19/02/2023	19/02/2023
					CT983	Surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	19/02/2023	19/02/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	19/02/2023	19/02/2023
12657	2023	All other reportable incidents	35852	Single-chamber implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
									A1102	Application Program Problem		
									A23	Use of Device Problem		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
									A1102	Application Program Problem		
									A23	Use of Device Problem		
12658	2023	All other reportable incidents	36241	Pacemaker/defibrillator lead	CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A072201	High Impedance		
							F26	No Health Consequences or Impact	A072201	High Impedance		
12659	2023	All other reportable incidents	47263	Cardiac resynchronization therapy implantable pacemaker	CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A0908	Incorrect, Inadequate or Imprecise Result or Readings	17/03/2023	17/03/2023
							F08	Hospitalization or Prolonged Hospitalization	A0908	Incorrect, Inadequate or Imprecise Result or Readings	17/03/2023	17/03/2023
							F1905	Device Revision or Replacement	A0908	Incorrect, Inadequate or Imprecise Result or Readings	17/03/2023	17/03/2023
					CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A0908	Incorrect, Inadequate or Imprecise Result or Readings	17/03/2023	17/03/2023
							F08	Hospitalization or Prolonged Hospitalization	A0908	Incorrect, Inadequate or Imprecise Result or Readings	17/03/2023	17/03/2023
							F1905	Device Revision or Replacement	A0908	Incorrect, Inadequate or Imprecise Result or Readings	17/03/2023	17/03/2023
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A0908	Incorrect, Inadequate or Imprecise Result or Readings	17/03/2023	17/03/2023
							F08	Hospitalization or Prolonged Hospitalization	A0908	Incorrect, Inadequate or Imprecise Result or Readings	17/03/2023	17/03/2023
							F1905	Device Revision or Replacement	A0908	Incorrect, Inadequate or Imprecise Result or Readings	17/03/2023	17/03/2023
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A0908	Incorrect, Inadequate or Imprecise Result or Readings	17/03/2023	17/03/2023
							F08	Hospitalization or Prolonged Hospitalization	A0908	Incorrect, Inadequate or Imprecise Result or Readings	17/03/2023	17/03/2023
							F1905	Device Revision or Replacement	A0908	Incorrect, Inadequate or Imprecise Result or Readings	17/03/2023	17/03/2023
12660	2023		47265	Dual-chamber implantable pacemaker, rate-responsive	CT141	Electrodes/Leads and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	19/02/2023	19/02/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	19/02/2023	19/02/2023
					CT145	Cardiovascular devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	19/02/2023	19/02/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	19/02/2023	19/02/2023
					CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	19/02/2023	19/02/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	19/02/2023	19/02/2023
					CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	19/02/2023	19/02/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	19/02/2023	19/02/2023
					CT233	Surgical	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	19/02/2023	19/02/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	19/02/2023	19/02/2023
					CT2406	Implantable	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	19/02/2023	19/02/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	19/02/2023	19/02/2023
					CT2407	Active implantable	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	19/02/2023	19/02/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	19/02/2023	19/02/2023
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	19/02/2023	19/02/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	19/02/2023	19/02/2023
					CT2727	Active implant devices/ancillaries	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	19/02/2023	19/02/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	19/02/2023	19/02/2023
					CT2750	Implantable cardiac electrical stimulators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	19/02/2023	19/02/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	19/02/2023	19/02/2023
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	19/02/2023	19/02/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	19/02/2023	19/02/2023
					CT2985	Active	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	19/02/2023	19/02/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	19/02/2023	19/02/2023
					CT3001	Implantable metal	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	19/02/2023	19/02/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	19/02/2023	19/02/2023
					CT3002	Implantable synthetic polymer	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	19/02/2023	19/02/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	19/02/2023	19/02/2023
					CT3006	Implantable inorganic material	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	19/02/2023	19/02/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	19/02/2023	19/02/2023
					CT314	Electrically powered	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	19/02/2023	19/02/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	19/02/2023	19/02/2023
					CT321	Long-term surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	19/02/2023	19/02/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	19/02/2023	19/02/2023
					CT449	Generators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	19/02/2023	19/02/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	19/02/2023	19/02/2023
					CT981	Single use	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	19/02/2023	19/02/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	19/02/2023	19/02/2023
					CT983	Surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	19/02/2023	19/02/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	19/02/2023	19/02/2023
12661	2023	All other reportable incidents	17882	Physiologic-monitoring defibrillation system	CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A07	Electrical /Electronic Property Problem		
							F27	Problem Identified During Non-Clinical Procedure	A07	Electrical /Electronic Property Problem		
12662	2023	All other reportable incidents	17882	Physiologic-monitoring defibrillation system	CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A071204	No Pacing		
							F26	No Health Consequences or Impact	A071204	No Pacing		
12663	2023	All other reportable incidents	47265	Dual-chamber implantable pacemaker, rate-responsive	CT141	Electrodes/Leads and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A071102	Failure to Interrogate		
							F1901	Additional Surgery	A071102	Failure to Interrogate		
					CT145	Cardiovascular devices	E2403	No Clinical Signs, Symptoms or Conditions	A071102	Failure to Interrogate		
							F1901	Additional Surgery	A071102	Failure to Interrogate		
					CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A071102	Failure to Interrogate		

						F1901	Additional Surgery	A071102	Failure to Interrogate			
				CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A071102	Failure to Interrogate			
						F1901	Additional Surgery	A071102	Failure to Interrogate			
				CT233	Surgical	E2403	No Clinical Signs, Symptoms or Conditions	A071102	Failure to Interrogate			
						F1901	Additional Surgery	A071102	Failure to Interrogate			
				CT2406	Implantable	E2403	No Clinical Signs, Symptoms or Conditions	A071102	Failure to Interrogate			
						F1901	Additional Surgery	A071102	Failure to Interrogate			
				CT2407	Active implantable	E2403	No Clinical Signs, Symptoms or Conditions	A071102	Failure to Interrogate			
						F1901	Additional Surgery	A071102	Failure to Interrogate			
				CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A071102	Failure to Interrogate			
						F1901	Additional Surgery	A071102	Failure to Interrogate			
				CT2727	Active implant devices/ancillaries	E2403	No Clinical Signs, Symptoms or Conditions	A071102	Failure to Interrogate			
						F1901	Additional Surgery	A071102	Failure to Interrogate			
				CT2750	Implantable cardiac electrical stimulators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A071102	Failure to Interrogate			
						F1901	Additional Surgery	A071102	Failure to Interrogate			
				CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A071102	Failure to Interrogate			
						F1901	Additional Surgery	A071102	Failure to Interrogate			
				CT2985	Active	E2403	No Clinical Signs, Symptoms or Conditions	A071102	Failure to Interrogate			
						F1901	Additional Surgery	A071102	Failure to Interrogate			
				CT3001	Implantable metal	E2403	No Clinical Signs, Symptoms or Conditions	A071102	Failure to Interrogate			
						F1901	Additional Surgery	A071102	Failure to Interrogate			
				CT3002	Implantable synthetic polymer	E2403	No Clinical Signs, Symptoms or Conditions	A071102	Failure to Interrogate			
						F1901	Additional Surgery	A071102	Failure to Interrogate			
				CT3006	Implantable inorganic material	E2403	No Clinical Signs, Symptoms or Conditions	A071102	Failure to Interrogate			
						F1901	Additional Surgery	A071102	Failure to Interrogate			
				CT314	Electrically powered	E2403	No Clinical Signs, Symptoms or Conditions	A071102	Failure to Interrogate			
						F1901	Additional Surgery	A071102	Failure to Interrogate			
				CT321	Long-term surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A071102	Failure to Interrogate			
						F1901	Additional Surgery	A071102	Failure to Interrogate			
				CT449	Generators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A071102	Failure to Interrogate			
						F1901	Additional Surgery	A071102	Failure to Interrogate			
				CT981	Single use	E2403	No Clinical Signs, Symptoms or Conditions	A071102	Failure to Interrogate			
						F1901	Additional Surgery	A071102	Failure to Interrogate			
				CT983	Surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A071102	Failure to Interrogate			
						F1901	Additional Surgery	A071102	Failure to Interrogate			
12664	2023		47265	Dual-chamber implantable pacemaker, rate-responsive	CT141	Electrodes/Leads and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem		
							F1901	Additional Surgery	A24	No Apparent Device Problem		
					CT145	Cardiovascular devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem		
							F1901	Additional Surgery	A24	No Apparent Device Problem		
					CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem		
							F1901	Additional Surgery	A24	No Apparent Device Problem		
					CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem		
							F1901	Additional Surgery	A24	No Apparent Device Problem		
					CT233	Surgical	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem		
							F1901	Additional Surgery	A24	No Apparent Device Problem		
					CT2406	Implantable	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem		
							F1901	Additional Surgery	A24	No Apparent Device Problem		
					CT2407	Active implantable	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem		
							F1901	Additional Surgery	A24	No Apparent Device Problem		
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem		
							F1901	Additional Surgery	A24	No Apparent Device Problem		
					CT2727	Active implant devices/ancillaries	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem		
							F1901	Additional Surgery	A24	No Apparent Device Problem		
					CT2750	Implantable cardiac electrical stimulators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem		
							F1901	Additional Surgery	A24	No Apparent Device Problem		
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem		
							F1901	Additional Surgery	A24	No Apparent Device Problem		
					CT2985	Active	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem		
							F1901	Additional Surgery	A24	No Apparent Device Problem		
					CT3001	Implantable metal	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem		
							F1901	Additional Surgery	A24	No Apparent Device Problem		
					CT3002	Implantable synthetic polymer	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem		
							F1901	Additional Surgery	A24	No Apparent Device Problem		
					CT3006	Implantable inorganic material	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem		
							F1901	Additional Surgery	A24	No Apparent Device Problem		
					CT314	Electrically powered	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem		
							F1901	Additional Surgery	A24	No Apparent Device Problem		
					CT321	Long-term surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem		
							F1901	Additional Surgery	A24	No Apparent Device Problem		
					CT449	Generators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem		
							F1901	Additional Surgery	A24	No Apparent Device Problem		
					CT981	Single use	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem		
							F1901	Additional Surgery	A24	No Apparent Device Problem		
					CT983	Surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem		
							F1901	Additional Surgery	A24	No Apparent Device Problem		
12665	2023		47265	Dual-chamber implantable pacemaker, rate-responsive	CT141	Electrodes/Leads and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem		
							F1901	Additional Surgery	A24	No Apparent Device Problem		
					CT145	Cardiovascular devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem		
							F1901	Additional Surgery	A24	No Apparent Device Problem		
					CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem		
							F1901	Additional Surgery	A24	No Apparent Device Problem		
					CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem		
							F1901	Additional Surgery	A24	No Apparent Device Problem		
					CT233	Surgical	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem		
							F1901	Additional Surgery	A24	No Apparent Device Problem		
					CT2406	Implantable	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem		
							F1901	Additional Surgery	A24	No Apparent Device Problem		
					CT2407	Active implantable	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem		

							F1901	Additional Surgery	A24	No Apparent Device Problem		
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem		
							F1901	Additional Surgery	A24	No Apparent Device Problem		
					CT2727	Active implant devices/ancillaries	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem		
							F1901	Additional Surgery	A24	No Apparent Device Problem		
					CT2750	Implantable cardiac electrical stimulators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem		
							F1901	Additional Surgery	A24	No Apparent Device Problem		
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem		
							F1901	Additional Surgery	A24	No Apparent Device Problem		
					CT2985	Active	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem		
							F1901	Additional Surgery	A24	No Apparent Device Problem		
					CT3001	Implantable metal	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem		
							F1901	Additional Surgery	A24	No Apparent Device Problem		
					CT3002	Implantable synthetic polymer	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem		
							F1901	Additional Surgery	A24	No Apparent Device Problem		
					CT3006	Implantable inorganic material	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem		
							F1901	Additional Surgery	A24	No Apparent Device Problem		
					CT314	Electrically powered	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem		
							F1901	Additional Surgery	A24	No Apparent Device Problem		
					CT321	Long-term surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem		
							F1901	Additional Surgery	A24	No Apparent Device Problem		
					CT449	Generators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem		
							F1901	Additional Surgery	A24	No Apparent Device Problem		
					CT981	Single use	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem		
							F1901	Additional Surgery	A24	No Apparent Device Problem		
					CT983	Surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem		
							F1901	Additional Surgery	A24	No Apparent Device Problem		
12666	2023		47265	Dual-chamber implantable pacemaker, rate-responsive	CT141	Electrodes/Leads and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem		
							F1901	Additional Surgery	A24	No Apparent Device Problem		
					CT145	Cardiovascular devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem		
							F1901	Additional Surgery	A24	No Apparent Device Problem		
					CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem		
							F1901	Additional Surgery	A24	No Apparent Device Problem		
					CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem		
							F1901	Additional Surgery	A24	No Apparent Device Problem		
					CT233	Surgical	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem		
							F1901	Additional Surgery	A24	No Apparent Device Problem		
					CT2406	Implantable	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem		
							F1901	Additional Surgery	A24	No Apparent Device Problem		
					CT2407	Active implantable	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem		
							F1901	Additional Surgery	A24	No Apparent Device Problem		
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem		
							F1901	Additional Surgery	A24	No Apparent Device Problem		
					CT2727	Active implant devices/ancillaries	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem		
							F1901	Additional Surgery	A24	No Apparent Device Problem		
					CT2750	Implantable cardiac electrical stimulators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem		
							F1901	Additional Surgery	A24	No Apparent Device Problem		
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem		
							F1901	Additional Surgery	A24	No Apparent Device Problem		
					CT2985	Active	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem		
							F1901	Additional Surgery	A24	No Apparent Device Problem		
					CT3001	Implantable metal	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem		
							F1901	Additional Surgery	A24	No Apparent Device Problem		
					CT3002	Implantable synthetic polymer	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem		
							F1901	Additional Surgery	A24	No Apparent Device Problem		
					CT3006	Implantable inorganic material	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem		
							F1901	Additional Surgery	A24	No Apparent Device Problem		
					CT314	Electrically powered	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem		
							F1901	Additional Surgery	A24	No Apparent Device Problem		
					CT321	Long-term surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem		
							F1901	Additional Surgery	A24	No Apparent Device Problem		
					CT449	Generators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem		
							F1901	Additional Surgery	A24	No Apparent Device Problem		
					CT981	Single use	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem		
							F1901	Additional Surgery	A24	No Apparent Device Problem		
					CT983	Surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem		
							F1901	Additional Surgery	A24	No Apparent Device Problem		
12667	2023	All other reportable incidents	47265	Dual-chamber implantable pacemaker, rate-responsive	CT141	Electrodes/Leads and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem		
							F1901	Additional Surgery	A24	No Apparent Device Problem		
					CT145	Cardiovascular devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem		
							F1901	Additional Surgery	A24	No Apparent Device Problem		
					CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem		
							F1901	Additional Surgery	A24	No Apparent Device Problem		
					CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem		
							F1901	Additional Surgery	A24	No Apparent Device Problem		
					CT233	Surgical	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem		
							F1901	Additional Surgery	A24	No Apparent Device Problem		
					CT2406	Implantable	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem		
							F1901	Additional Surgery	A24	No Apparent Device Problem		
					CT2407	Active implantable	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem		
							F1901	Additional Surgery	A24	No Apparent Device Problem		
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem		
							F1901	Additional Surgery	A24	No Apparent Device Problem		
					CT2727	Active implant devices/ancillaries	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem		
							F1901	Additional Surgery	A24	No Apparent Device Problem		
					CT2750	Implantable cardiac electrical stimulators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem		
							F1901	Additional Surgery	A24	No Apparent Device Problem		
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem		

							F1901	Additional Surgery	A24	No Apparent Device Problem		
					CT2985	Active	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem		
							F1901	Additional Surgery	A24	No Apparent Device Problem		
					CT3001	Implantable metal	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem		
							F1901	Additional Surgery	A24	No Apparent Device Problem		
					CT3002	Implantable synthetic polymer	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem		
							F1901	Additional Surgery	A24	No Apparent Device Problem		
					CT3006	Implantable inorganic material	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem		
							F1901	Additional Surgery	A24	No Apparent Device Problem		
					CT314	Electrically powered	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem		
							F1901	Additional Surgery	A24	No Apparent Device Problem		
					CT321	Long-term surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem		
							F1901	Additional Surgery	A24	No Apparent Device Problem		
					CT449	Generators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem		
							F1901	Additional Surgery	A24	No Apparent Device Problem		
					CT981	Single use	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem		
							F1901	Additional Surgery	A24	No Apparent Device Problem		
					CT983	Surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem		
							F1901	Additional Surgery	A24	No Apparent Device Problem		
12668	2023		47265	Dual-chamber implantable pacemaker, rate-responsive	CT141	Electrodes/Leads and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem		
							F1901	Additional Surgery	A24	No Apparent Device Problem		
					CT145	Cardiovascular devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem		
							F1901	Additional Surgery	A24	No Apparent Device Problem		
					CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem		
							F1901	Additional Surgery	A24	No Apparent Device Problem		
					CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem		
							F1901	Additional Surgery	A24	No Apparent Device Problem		
					CT233	Surgical	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem		
							F1901	Additional Surgery	A24	No Apparent Device Problem		
					CT2406	Implantable	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem		
							F1901	Additional Surgery	A24	No Apparent Device Problem		
					CT2407	Active implantable	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem		
							F1901	Additional Surgery	A24	No Apparent Device Problem		
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem		
							F1901	Additional Surgery	A24	No Apparent Device Problem		
					CT2727	Active implant devices/ancillaries	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem		
							F1901	Additional Surgery	A24	No Apparent Device Problem		
					CT2750	Implantable cardiac electrical stimulators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem		
							F1901	Additional Surgery	A24	No Apparent Device Problem		
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem		
							F1901	Additional Surgery	A24	No Apparent Device Problem		
					CT2985	Active	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem		
							F1901	Additional Surgery	A24	No Apparent Device Problem		
					CT3001	Implantable metal	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem		
							F1901	Additional Surgery	A24	No Apparent Device Problem		
					CT3002	Implantable synthetic polymer	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem		
							F1901	Additional Surgery	A24	No Apparent Device Problem		
					CT3006	Implantable inorganic material	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem		
							F1901	Additional Surgery	A24	No Apparent Device Problem		
					CT314	Electrically powered	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem		
							F1901	Additional Surgery	A24	No Apparent Device Problem		
					CT321	Long-term surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem		
							F1901	Additional Surgery	A24	No Apparent Device Problem		
					CT449	Generators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem		
							F1901	Additional Surgery	A24	No Apparent Device Problem		
					CT981	Single use	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem		
							F1901	Additional Surgery	A24	No Apparent Device Problem		
					CT983	Surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem		
							F1901	Additional Surgery	A24	No Apparent Device Problem		
12669	2023	All other reportable incidents	47263	Cardiac resynchronization therapy implantable pacemaker	CT1685	Cardiac pacemakers	E011903	Syncope/Fainting	A0203	Defective Device	14/03/2023	14/03/2023
									A070101	Failure to Capture	14/03/2023	14/03/2023
									A0908	Incorrect, Inadequate or Imprecise Result or Readings	14/03/2023	14/03/2023
							F08	Hospitalization or Prolonged Hospitalization	A0203	Defective Device	14/03/2023	14/03/2023
									A070101	Failure to Capture	14/03/2023	14/03/2023
									A0908	Incorrect, Inadequate or Imprecise Result or Readings	14/03/2023	14/03/2023
							F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A0203	Defective Device	14/03/2023	14/03/2023
									A070101	Failure to Capture	14/03/2023	14/03/2023
									A0908	Incorrect, Inadequate or Imprecise Result or Readings	14/03/2023	14/03/2023
							F1905	Device Revision or Replacement	A0203	Defective Device	14/03/2023	14/03/2023
									A070101	Failure to Capture	14/03/2023	14/03/2023
									A0908	Incorrect, Inadequate or Imprecise Result or Readings	14/03/2023	14/03/2023
					CT257	Implantable cardiac pacemakers	E011903	Syncope/Fainting	A0203	Defective Device	14/03/2023	14/03/2023
									A070101	Failure to Capture	14/03/2023	14/03/2023
									A0908	Incorrect, Inadequate or Imprecise Result or Readings	14/03/2023	14/03/2023
							F08	Hospitalization or Prolonged Hospitalization	A0203	Defective Device	14/03/2023	14/03/2023
									A070101	Failure to Capture	14/03/2023	14/03/2023
									A0908	Incorrect, Inadequate or Imprecise Result or Readings	14/03/2023	14/03/2023
							F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A0203	Defective Device	14/03/2023	14/03/2023
									A070101	Failure to Capture	14/03/2023	14/03/2023
									A0908	Incorrect, Inadequate or Imprecise Result or Readings	14/03/2023	14/03/2023
							F1905	Device Revision or Replacement	A0203	Defective Device	14/03/2023	14/03/2023
									A070101	Failure to Capture	14/03/2023	14/03/2023

								A0908	Incorrect, Inadequate or Imprecise Result or Readings	14/03/2023	14/03/2023
				CT2866	Implantable cardiac electrical stimulators	E011903	Syncope/Fainting	A0203	Defective Device	14/03/2023	14/03/2023
								A070101	Failure to Capture	14/03/2023	14/03/2023
								A0908	Incorrect, Inadequate or Imprecise Result or Readings	14/03/2023	14/03/2023
						F08	Hospitalization or Prolonged Hospitalization	A0203	Defective Device	14/03/2023	14/03/2023
								A070101	Failure to Capture	14/03/2023	14/03/2023
								A0908	Incorrect, Inadequate or Imprecise Result or Readings	14/03/2023	14/03/2023
						F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A0203	Defective Device	14/03/2023	14/03/2023
								A070101	Failure to Capture	14/03/2023	14/03/2023
								A0908	Incorrect, Inadequate or Imprecise Result or Readings	14/03/2023	14/03/2023
						F1905	Device Revision or Replacement	A0203	Defective Device	14/03/2023	14/03/2023
								A070101	Failure to Capture	14/03/2023	14/03/2023
								A0908	Incorrect, Inadequate or Imprecise Result or Readings	14/03/2023	14/03/2023
12670	2023	All other reportable incidents	36241	Pacemaker/defibrillator lead	CT1684	Cardiac pacemakers and associated devices	E2104	Electric Shock	A0203	Defective Device		
									A070909	Oversensing		
									A071302	Inappropriate/Inadequate Shock/Stimulation		
									A090801	Signal Artifact/Noise		
						F08	Hospitalization or Prolonged Hospitalization	A0203	Defective Device		
									A070909	Oversensing		
									A071302	Inappropriate/Inadequate Shock/Stimulation		
									A090801	Signal Artifact/Noise		
						F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A0203	Defective Device		
									A070909	Oversensing		
									A071302	Inappropriate/Inadequate Shock/Stimulation		
									A090801	Signal Artifact/Noise		
12671	2023	All other reportable incidents	17882	Physiologic-monitoring defibrillation system	CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A071301	Failure to Deliver Shock/Stimulation		
							F26	No Health Consequences or Impact	A071301	Failure to Deliver Shock/Stimulation		
12672	2023	All other reportable incidents	47265	Dual-chamber implantable pacemaker, rate-responsive	CT141	Electrodes/Leads and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A072201	High Impedance		
							F26	No Health Consequences or Impact	A072201	High Impedance		
					CT145	Cardiovascular devices	E2403	No Clinical Signs, Symptoms or Conditions	A072201	High Impedance		
							F26	No Health Consequences or Impact	A072201	High Impedance		
					CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A072201	High Impedance		
							F26	No Health Consequences or Impact	A072201	High Impedance		
					CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A072201	High Impedance		
							F26	No Health Consequences or Impact	A072201	High Impedance		
					CT233	Surgical	E2403	No Clinical Signs, Symptoms or Conditions	A072201	High Impedance		
							F26	No Health Consequences or Impact	A072201	High Impedance		
					CT2406	Implantable	E2403	No Clinical Signs, Symptoms or Conditions	A072201	High Impedance		
							F26	No Health Consequences or Impact	A072201	High Impedance		
					CT2407	Active implantable	E2403	No Clinical Signs, Symptoms or Conditions	A072201	High Impedance		
							F26	No Health Consequences or Impact	A072201	High Impedance		
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A072201	High Impedance		
							F26	No Health Consequences or Impact	A072201	High Impedance		
					CT2727	Active implant devices/ancillaries	E2403	No Clinical Signs, Symptoms or Conditions	A072201	High Impedance		
							F26	No Health Consequences or Impact	A072201	High Impedance		
					CT2750	Implantable cardiac electrical stimulators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A072201	High Impedance		
							F26	No Health Consequences or Impact	A072201	High Impedance		
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A072201	High Impedance		
							F26	No Health Consequences or Impact	A072201	High Impedance		
					CT2985	Active	E2403	No Clinical Signs, Symptoms or Conditions	A072201	High Impedance		
							F26	No Health Consequences or Impact	A072201	High Impedance		
					CT3001	Implantable metal	E2403	No Clinical Signs, Symptoms or Conditions	A072201	High Impedance		
							F26	No Health Consequences or Impact	A072201	High Impedance		
					CT3002	Implantable synthetic polymer	E2403	No Clinical Signs, Symptoms or Conditions	A072201	High Impedance		
							F26	No Health Consequences or Impact	A072201	High Impedance		
					CT3006	Implantable inorganic material	E2403	No Clinical Signs, Symptoms or Conditions	A072201	High Impedance		
							F26	No Health Consequences or Impact	A072201	High Impedance		
					CT314	Electrically powered	E2403	No Clinical Signs, Symptoms or Conditions	A072201	High Impedance		
							F26	No Health Consequences or Impact	A072201	High Impedance		
					CT321	Long-term surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A072201	High Impedance		
							F26	No Health Consequences or Impact	A072201	High Impedance		
					CT449	Generators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A072201	High Impedance		
							F26	No Health Consequences or Impact	A072201	High Impedance		
					CT981	Single use	E2403	No Clinical Signs, Symptoms or Conditions	A072201	High Impedance		
							F26	No Health Consequences or Impact	A072201	High Impedance		
					CT983	Surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A072201	High Impedance		
							F26	No Health Consequences or Impact	A072201	High Impedance		
12673	2023	All other reportable incidents	36241	Pacemaker/defibrillator lead	CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070909	Oversensing		
									A0722	Impedance Problem		
									A090801	Signal Artifact/Noise		
							F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A070909	Oversensing		
									A0722	Impedance Problem		
									A090801	Signal Artifact/Noise		
12674	2023	All other reportable incidents	35223	Endocardial/interventricular septal pacing lead	CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070102	High Capture Threshold		
									A072201	High Impedance		
							F23	Unexpected Medical Intervention	A070102	High Capture Threshold		
									A072201	High Impedance		
12675	2023	All other reportable incidents	17882	Physiologic-monitoring defibrillation system	CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A0719	Unexpected Shutdown		
							F27	Problem Identified During Non-Clinical Procedure	A0719	Unexpected Shutdown		
12676	2023	All other reportable incidents	35852	Single-chamber implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A072201	High Impedance		
							F08	Hospitalization or Prolonged Hospitalization	A072201	High Impedance		
							F2203	Imaging Required	A072201	High Impedance		
12677	2023	All other reportable incidents	35852	Single-chamber implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E1705	Burning Sensation	A0709	Device Sensing Problem		
									A070909	Oversensing		

									A071302	Inappropriate/Inadequate Shock/Stimulation		
									A12	Connection Problem		
							E2104	Electric Shock	A0709	Device Sensing Problem		
									A070909	Oversensing		
									A071302	Inappropriate/Inadequate Shock/Stimulation		
									A12	Connection Problem		
							F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A0709	Device Sensing Problem		
									A070909	Oversensing		
									A071302	Inappropriate/Inadequate Shock/Stimulation		
									A12	Connection Problem		
							F2203	Imaging Required	A0709	Device Sensing Problem		
									A070909	Oversensing		
									A071302	Inappropriate/Inadequate Shock/Stimulation		
									A12	Connection Problem		
12678	2023	All other reportable incidents	35223	Endocardial/interventricular septal pacing lead	CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070101	Failure to Capture	09/03/2023	09/03/2023
									A072201	High Impedance	09/03/2023	09/03/2023
							F1901	Additional Surgery	A070101	Failure to Capture	09/03/2023	09/03/2023
									A072201	High Impedance	09/03/2023	09/03/2023
12679	2023		47265	Dual-chamber implantable pacemaker, rate-responsive	CT141	Electrodes/Leads and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	10/03/2023	10/03/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	10/03/2023	10/03/2023
					CT145	Cardiovascular devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	10/03/2023	10/03/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	10/03/2023	10/03/2023
					CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	10/03/2023	10/03/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	10/03/2023	10/03/2023
					CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	10/03/2023	10/03/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	10/03/2023	10/03/2023
					CT233	Surgical	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	10/03/2023	10/03/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	10/03/2023	10/03/2023
					CT2406	Implantable	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	10/03/2023	10/03/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	10/03/2023	10/03/2023
					CT2407	Active implantable	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	10/03/2023	10/03/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	10/03/2023	10/03/2023
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	10/03/2023	10/03/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	10/03/2023	10/03/2023
					CT2727	Active implant devices/ancillaries	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	10/03/2023	10/03/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	10/03/2023	10/03/2023
					CT2750	Implantable cardiac electrical stimulators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	10/03/2023	10/03/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	10/03/2023	10/03/2023
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	10/03/2023	10/03/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	10/03/2023	10/03/2023
					CT2985	Active	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	10/03/2023	10/03/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	10/03/2023	10/03/2023
					CT3001	Implantable metal	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	10/03/2023	10/03/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	10/03/2023	10/03/2023
					CT3002	Implantable synthetic polymer	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	10/03/2023	10/03/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	10/03/2023	10/03/2023
					CT3006	Implantable inorganic material	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	10/03/2023	10/03/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	10/03/2023	10/03/2023
					CT314	Electrically powered	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	10/03/2023	10/03/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	10/03/2023	10/03/2023
					CT321	Long-term surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	10/03/2023	10/03/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	10/03/2023	10/03/2023
					CT449	Generators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	10/03/2023	10/03/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	10/03/2023	10/03/2023
					CT981	Single use	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	10/03/2023	10/03/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	10/03/2023	10/03/2023
					CT983	Surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	10/03/2023	10/03/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	10/03/2023	10/03/2023
12680	2023	All other reportable incidents	36241	Pacemaker/defibrillator lead	CT1684	Cardiac pacemakers and associated devices	E2104	Electric Shock	A070909	Oversensing		
									A071302	Inappropriate/Inadequate Shock/Stimulation		
							F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A070909	Oversensing		
									A071302	Inappropriate/Inadequate Shock/Stimulation		
12681	2023	All other reportable incidents	35852	Single-chamber implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
12682	2023		47265	Dual-chamber implantable pacemaker, rate-responsive	CT141	Electrodes/Leads and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	11/03/2023	11/03/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	11/03/2023	11/03/2023
					CT145	Cardiovascular devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	11/03/2023	11/03/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	11/03/2023	11/03/2023
					CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	11/03/2023	11/03/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	11/03/2023	11/03/2023
					CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	11/03/2023	11/03/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	11/03/2023	11/03/2023
					CT233	Surgical	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	11/03/2023	11/03/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	11/03/2023	11/03/2023
					CT2406	Implantable	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	11/03/2023	11/03/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	11/03/2023	11/03/2023
					CT2407	Active implantable	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	11/03/2023	11/03/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	11/03/2023	11/03/2023
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	11/03/2023	11/03/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	11/03/2023	11/03/2023
					CT2727	Active implant devices/ancillaries	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	11/03/2023	11/03/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	11/03/2023	11/03/2023
					CT2750	Implantable cardiac electrical stimulators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	11/03/2023	11/03/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	11/03/2023	11/03/2023
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	11/03/2023	11/03/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	11/03/2023	11/03/2023

					CT2985	Active	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	11/03/2023	11/03/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	11/03/2023	11/03/2023
					CT3001	Implantable metal	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	11/03/2023	11/03/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	11/03/2023	11/03/2023
					CT3002	Implantable synthetic polymer	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	11/03/2023	11/03/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	11/03/2023	11/03/2023
					CT3006	Implantable inorganic material	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	11/03/2023	11/03/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	11/03/2023	11/03/2023
					CT314	Electrically powered	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	11/03/2023	11/03/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	11/03/2023	11/03/2023
					CT321	Long-term surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	11/03/2023	11/03/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	11/03/2023	11/03/2023
					CT449	Generators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	11/03/2023	11/03/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	11/03/2023	11/03/2023
					CT981	Single use	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	11/03/2023	11/03/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	11/03/2023	11/03/2023
					CT983	Surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	11/03/2023	11/03/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	11/03/2023	11/03/2023
12683	2023	All other reportable incidents	35852	Single-chamber implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E2104	Electric Shock	A070909	Oversensing		
									A071302	Inappropriate/Inadequate Shock/Stimulation		
							F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A070909	Oversensing		
									A071302	Inappropriate/Inadequate Shock/Stimulation		
12684	2023	All other reportable incidents	35852	Single-chamber implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
12685	2023	All other reportable incidents	35039	Epicardial pacing lead	CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070102	High Capture Threshold	20/03/2023	20/03/2023
							F08	Hospitalization or Prolonged Hospitalization	A070102	High Capture Threshold	20/03/2023	20/03/2023
							F12	Serious Injury/ Illness/ Impairment	A070102	High Capture Threshold	20/03/2023	20/03/2023
							F1905	Device Revision or Replacement	A070102	High Capture Threshold	20/03/2023	20/03/2023
							F23	Unexpected Medical Intervention	A070102	High Capture Threshold	20/03/2023	20/03/2023
12686	2023	All other reportable incidents	35039	Epicardial pacing lead	CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070102	High Capture Threshold		
							F08	Hospitalization or Prolonged Hospitalization	A070102	High Capture Threshold		
							F12	Serious Injury/ Illness/ Impairment	A070102	High Capture Threshold		
							F1905	Device Revision or Replacement	A070102	High Capture Threshold		
							F23	Unexpected Medical Intervention	A070102	High Capture Threshold		
12687	2023	All other reportable incidents	35223	Endocardial/interventricular septal pacing lead	CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070909	Oversensing		
									A070910	Undersensing		
							F12	Serious Injury/ Illness/ Impairment	A070909	Oversensing		
									A070910	Undersensing		
							F23	Unexpected Medical Intervention	A070909	Oversensing		
									A070910	Undersensing		
12688	2023	All other reportable incidents	47265	Dual-chamber implantable pacemaker, rate-responsive	CT141	Electrodes/Leads and associated devices	E0601	Arrhythmia	A071102	Failure to Interrogate		
									A0712	Pacing Problem		
							F1908	Prolonged Surgery	A071102	Failure to Interrogate		
									A0712	Pacing Problem		
					CT145	Cardiovascular devices	E0601	Arrhythmia	A071102	Failure to Interrogate		
									A0712	Pacing Problem		
							F1908	Prolonged Surgery	A071102	Failure to Interrogate		
									A0712	Pacing Problem		
					CT1684	Cardiac pacemakers and associated devices	E0601	Arrhythmia	A071102	Failure to Interrogate		
									A0712	Pacing Problem		
							F1908	Prolonged Surgery	A071102	Failure to Interrogate		
									A0712	Pacing Problem		
					CT1685	Cardiac pacemakers	E0601	Arrhythmia	A071102	Failure to Interrogate		
									A0712	Pacing Problem		
							F1908	Prolonged Surgery	A071102	Failure to Interrogate		
									A0712	Pacing Problem		
					CT233	Surgical	E0601	Arrhythmia	A071102	Failure to Interrogate		
									A0712	Pacing Problem		
							F1908	Prolonged Surgery	A071102	Failure to Interrogate		
									A0712	Pacing Problem		
					CT2406	Implantable	E0601	Arrhythmia	A071102	Failure to Interrogate		
									A0712	Pacing Problem		
							F1908	Prolonged Surgery	A071102	Failure to Interrogate		
									A0712	Pacing Problem		
					CT2407	Active implantable	E0601	Arrhythmia	A071102	Failure to Interrogate		
									A0712	Pacing Problem		
							F1908	Prolonged Surgery	A071102	Failure to Interrogate		
									A0712	Pacing Problem		
					CT257	Implantable cardiac pacemakers	E0601	Arrhythmia	A071102	Failure to Interrogate		
									A0712	Pacing Problem		
							F1908	Prolonged Surgery	A071102	Failure to Interrogate		
									A0712	Pacing Problem		
					CT2727	Active implant devices/ancillaries	E0601	Arrhythmia	A071102	Failure to Interrogate		
									A0712	Pacing Problem		
							F1908	Prolonged Surgery	A071102	Failure to Interrogate		
									A0712	Pacing Problem		
					CT2750	Implantable cardiac electrical stimulators and associated devices	E0601	Arrhythmia	A071102	Failure to Interrogate		
									A0712	Pacing Problem		
							F1908	Prolonged Surgery	A071102	Failure to Interrogate		
									A0712	Pacing Problem		
					CT2866	Implantable cardiac electrical stimulators	E0601	Arrhythmia	A071102	Failure to Interrogate		
									A0712	Pacing Problem		
							F1908	Prolonged Surgery	A071102	Failure to Interrogate		
									A0712	Pacing Problem		
					CT2985	Active	E0601	Arrhythmia	A071102	Failure to Interrogate		
									A0712	Pacing Problem		
							F1908	Prolonged Surgery	A071102	Failure to Interrogate		
									A0712	Pacing Problem		

					CT3001	Implantable metal	E0601	Arrhythmia	A071102	Failure to Interrogate		
									A0712	Pacing Problem		
							F1908	Prolonged Surgery	A071102	Failure to Interrogate		
									A0712	Pacing Problem		
					CT3002	Implantable synthetic polymer	E0601	Arrhythmia	A071102	Failure to Interrogate		
									A0712	Pacing Problem		
							F1908	Prolonged Surgery	A071102	Failure to Interrogate		
									A0712	Pacing Problem		
					CT3006	Implantable inorganic material	E0601	Arrhythmia	A071102	Failure to Interrogate		
									A0712	Pacing Problem		
							F1908	Prolonged Surgery	A071102	Failure to Interrogate		
									A0712	Pacing Problem		
					CT314	Electrically powered	E0601	Arrhythmia	A071102	Failure to Interrogate		
									A0712	Pacing Problem		
							F1908	Prolonged Surgery	A071102	Failure to Interrogate		
									A0712	Pacing Problem		
					CT321	Long-term surgical invasive	E0601	Arrhythmia	A071102	Failure to Interrogate		
									A0712	Pacing Problem		
							F1908	Prolonged Surgery	A071102	Failure to Interrogate		
									A0712	Pacing Problem		
					CT449	Generators and associated devices	E0601	Arrhythmia	A071102	Failure to Interrogate		
									A0712	Pacing Problem		
							F1908	Prolonged Surgery	A071102	Failure to Interrogate		
									A0712	Pacing Problem		
					CT981	Single use	E0601	Arrhythmia	A071102	Failure to Interrogate		
									A0712	Pacing Problem		
							F1908	Prolonged Surgery	A071102	Failure to Interrogate		
									A0712	Pacing Problem		
					CT983	Surgical invasive	E0601	Arrhythmia	A071102	Failure to Interrogate		
									A0712	Pacing Problem		
							F1908	Prolonged Surgery	A071102	Failure to Interrogate		
									A0712	Pacing Problem		
12689	2023	All other reportable incidents	47263	Cardiac resynchronization therapy implantable pacemaker	CT1684	Cardiac pacemakers and associated devices	E0112	Dizziness	A070909	Oversensing	09/03/2023	09/03/2023
									A0712	Pacing Problem	09/03/2023	09/03/2023
									A0908	Incorrect, Inadequate or Imprecise Result or Readings	09/03/2023	09/03/2023
									A090801	Signal Artifact/Noise	09/03/2023	09/03/2023
							E011903	Syncope/Fainting	A070909	Oversensing	09/03/2023	09/03/2023
									A0712	Pacing Problem	09/03/2023	09/03/2023
									A0908	Incorrect, Inadequate or Imprecise Result or Readings	09/03/2023	09/03/2023
									A090801	Signal Artifact/Noise	09/03/2023	09/03/2023
							E2007	Fall	A070909	Oversensing	09/03/2023	09/03/2023
									A0712	Pacing Problem	09/03/2023	09/03/2023
									A0908	Incorrect, Inadequate or Imprecise Result or Readings	09/03/2023	09/03/2023
									A090801	Signal Artifact/Noise	09/03/2023	09/03/2023
							E2103	Device Overstimulation of Tissue	A070909	Oversensing	09/03/2023	09/03/2023
									A0712	Pacing Problem	09/03/2023	09/03/2023
									A0908	Incorrect, Inadequate or Imprecise Result or Readings	09/03/2023	09/03/2023
									A090801	Signal Artifact/Noise	09/03/2023	09/03/2023
							F08	Hospitalization or Prolonged Hospitalization	A070909	Oversensing	09/03/2023	09/03/2023
									A0712	Pacing Problem	09/03/2023	09/03/2023
									A0908	Incorrect, Inadequate or Imprecise Result or Readings	09/03/2023	09/03/2023
									A090801	Signal Artifact/Noise	09/03/2023	09/03/2023
							F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A070909	Oversensing	09/03/2023	09/03/2023
									A0712	Pacing Problem	09/03/2023	09/03/2023
									A0908	Incorrect, Inadequate or Imprecise Result or Readings	09/03/2023	09/03/2023
									A090801	Signal Artifact/Noise	09/03/2023	09/03/2023
							F1905	Device Revision or Replacement	A070909	Oversensing	09/03/2023	09/03/2023
									A0712	Pacing Problem	09/03/2023	09/03/2023
									A0908	Incorrect, Inadequate or Imprecise Result or Readings	09/03/2023	09/03/2023
									A090801	Signal Artifact/Noise	09/03/2023	09/03/2023
					CT1685	Cardiac pacemakers	E0112	Dizziness	A070909	Oversensing	09/03/2023	09/03/2023
									A0712	Pacing Problem	09/03/2023	09/03/2023
									A0908	Incorrect, Inadequate or Imprecise Result or Readings	09/03/2023	09/03/2023
									A090801	Signal Artifact/Noise	09/03/2023	09/03/2023
							E011903	Syncope/Fainting	A070909	Oversensing	09/03/2023	09/03/2023
									A0712	Pacing Problem	09/03/2023	09/03/2023
									A0908	Incorrect, Inadequate or Imprecise Result or Readings	09/03/2023	09/03/2023
									A090801	Signal Artifact/Noise	09/03/2023	09/03/2023
							E2007	Fall	A070909	Oversensing	09/03/2023	09/03/2023
									A0712	Pacing Problem	09/03/2023	09/03/2023
									A0908	Incorrect, Inadequate or Imprecise Result or Readings	09/03/2023	09/03/2023
									A090801	Signal Artifact/Noise	09/03/2023	09/03/2023
							E2103	Device Overstimulation of Tissue	A070909	Oversensing	09/03/2023	09/03/2023
									A0712	Pacing Problem	09/03/2023	09/03/2023
									A0908	Incorrect, Inadequate or Imprecise Result or Readings	09/03/2023	09/03/2023
									A090801	Signal Artifact/Noise	09/03/2023	09/03/2023
							F08	Hospitalization or Prolonged Hospitalization	A070909	Oversensing	09/03/2023	09/03/2023
									A0712	Pacing Problem	09/03/2023	09/03/2023
									A0908	Incorrect, Inadequate or Imprecise Result or Readings	09/03/2023	09/03/2023

									A090801	Signal Artifact/Noise	09/03/2023	09/03/2023
							F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A070909	Oversensing	09/03/2023	09/03/2023
									A0712	Pacing Problem	09/03/2023	09/03/2023
									A0908	Incorrect, Inadequate or Imprecise Result or Readings	09/03/2023	09/03/2023
									A090801	Signal Artifact/Noise	09/03/2023	09/03/2023
							F1905	Device Revision or Replacement	A070909	Oversensing	09/03/2023	09/03/2023
									A0712	Pacing Problem	09/03/2023	09/03/2023
									A0908	Incorrect, Inadequate or Imprecise Result or Readings	09/03/2023	09/03/2023
									A090801	Signal Artifact/Noise	09/03/2023	09/03/2023
					CT257	Implantable cardiac pacemakers	E0112	Dizziness	A070909	Oversensing	09/03/2023	09/03/2023
									A0712	Pacing Problem	09/03/2023	09/03/2023
									A0908	Incorrect, Inadequate or Imprecise Result or Readings	09/03/2023	09/03/2023
									A090801	Signal Artifact/Noise	09/03/2023	09/03/2023
							E011903	Syncope/Fainting	A070909	Oversensing	09/03/2023	09/03/2023
									A0712	Pacing Problem	09/03/2023	09/03/2023
									A0908	Incorrect, Inadequate or Imprecise Result or Readings	09/03/2023	09/03/2023
									A090801	Signal Artifact/Noise	09/03/2023	09/03/2023
							E2007	Fall	A070909	Oversensing	09/03/2023	09/03/2023
									A0712	Pacing Problem	09/03/2023	09/03/2023
									A0908	Incorrect, Inadequate or Imprecise Result or Readings	09/03/2023	09/03/2023
									A090801	Signal Artifact/Noise	09/03/2023	09/03/2023
							E2103	Device Overstimulation of Tissue	A070909	Oversensing	09/03/2023	09/03/2023
									A0712	Pacing Problem	09/03/2023	09/03/2023
									A0908	Incorrect, Inadequate or Imprecise Result or Readings	09/03/2023	09/03/2023
									A090801	Signal Artifact/Noise	09/03/2023	09/03/2023
							F08	Hospitalization or Prolonged Hospitalization	A070909	Oversensing	09/03/2023	09/03/2023
									A0712	Pacing Problem	09/03/2023	09/03/2023
									A0908	Incorrect, Inadequate or Imprecise Result or Readings	09/03/2023	09/03/2023
									A090801	Signal Artifact/Noise	09/03/2023	09/03/2023
							F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A070909	Oversensing	09/03/2023	09/03/2023
									A0712	Pacing Problem	09/03/2023	09/03/2023
									A0908	Incorrect, Inadequate or Imprecise Result or Readings	09/03/2023	09/03/2023
									A090801	Signal Artifact/Noise	09/03/2023	09/03/2023
							F1905	Device Revision or Replacement	A070909	Oversensing	09/03/2023	09/03/2023
									A0712	Pacing Problem	09/03/2023	09/03/2023
									A0908	Incorrect, Inadequate or Imprecise Result or Readings	09/03/2023	09/03/2023
									A090801	Signal Artifact/Noise	09/03/2023	09/03/2023
					CT2866	Implantable cardiac electrical stimulators	E0112	Dizziness	A070909	Oversensing	09/03/2023	09/03/2023
									A0712	Pacing Problem	09/03/2023	09/03/2023
									A0908	Incorrect, Inadequate or Imprecise Result or Readings	09/03/2023	09/03/2023
									A090801	Signal Artifact/Noise	09/03/2023	09/03/2023
							E011903	Syncope/Fainting	A070909	Oversensing	09/03/2023	09/03/2023
									A0712	Pacing Problem	09/03/2023	09/03/2023
									A0908	Incorrect, Inadequate or Imprecise Result or Readings	09/03/2023	09/03/2023
									A090801	Signal Artifact/Noise	09/03/2023	09/03/2023
							E2007	Fall	A070909	Oversensing	09/03/2023	09/03/2023
									A0712	Pacing Problem	09/03/2023	09/03/2023
									A0908	Incorrect, Inadequate or Imprecise Result or Readings	09/03/2023	09/03/2023
									A090801	Signal Artifact/Noise	09/03/2023	09/03/2023
							E2103	Device Overstimulation of Tissue	A070909	Oversensing	09/03/2023	09/03/2023
									A0712	Pacing Problem	09/03/2023	09/03/2023
									A0908	Incorrect, Inadequate or Imprecise Result or Readings	09/03/2023	09/03/2023
									A090801	Signal Artifact/Noise	09/03/2023	09/03/2023
							F08	Hospitalization or Prolonged Hospitalization	A070909	Oversensing	09/03/2023	09/03/2023
									A0712	Pacing Problem	09/03/2023	09/03/2023
									A0908	Incorrect, Inadequate or Imprecise Result or Readings	09/03/2023	09/03/2023
									A090801	Signal Artifact/Noise	09/03/2023	09/03/2023
							F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A070909	Oversensing	09/03/2023	09/03/2023
									A0712	Pacing Problem	09/03/2023	09/03/2023
									A0908	Incorrect, Inadequate or Imprecise Result or Readings	09/03/2023	09/03/2023
									A090801	Signal Artifact/Noise	09/03/2023	09/03/2023
							F1905	Device Revision or Replacement	A070909	Oversensing	09/03/2023	09/03/2023
									A0712	Pacing Problem	09/03/2023	09/03/2023
									A0908	Incorrect, Inadequate or Imprecise Result or Readings	09/03/2023	09/03/2023
									A090801	Signal Artifact/Noise	09/03/2023	09/03/2023
12690	2023		47265	Dual-chamber implantable pacemaker, rate-responsive	CT141	Electrodes/Leads and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem		
							F1901	Additional Surgery	A24	No Apparent Device Problem		
					CT145	Cardiovascular devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem		
							F1901	Additional Surgery	A24	No Apparent Device Problem		
					CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem		
							F1901	Additional Surgery	A24	No Apparent Device Problem		
					CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem		
							F1901	Additional Surgery	A24	No Apparent Device Problem		
					CT233	Surgical	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem		
							F1901	Additional Surgery	A24	No Apparent Device Problem		
					CT2406	Implantable	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem		

							F1901	Additional Surgery	A24	No Apparent Device Problem		
					CT2407	Active implantable	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem		
							F1901	Additional Surgery	A24	No Apparent Device Problem		
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem		
							F1901	Additional Surgery	A24	No Apparent Device Problem		
					CT2727	Active implant devices/ancillaries	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem		
							F1901	Additional Surgery	A24	No Apparent Device Problem		
					CT2750	Implantable cardiac electrical stimulators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem		
							F1901	Additional Surgery	A24	No Apparent Device Problem		
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem		
							F1901	Additional Surgery	A24	No Apparent Device Problem		
					CT2985	Active	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem		
							F1901	Additional Surgery	A24	No Apparent Device Problem		
					CT3001	Implantable metal	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem		
							F1901	Additional Surgery	A24	No Apparent Device Problem		
					CT3002	Implantable synthetic polymer	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem		
							F1901	Additional Surgery	A24	No Apparent Device Problem		
					CT3006	Implantable inorganic material	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem		
							F1901	Additional Surgery	A24	No Apparent Device Problem		
					CT314	Electrically powered	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem		
							F1901	Additional Surgery	A24	No Apparent Device Problem		
					CT321	Long-term surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem		
							F1901	Additional Surgery	A24	No Apparent Device Problem		
					CT449	Generators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem		
							F1901	Additional Surgery	A24	No Apparent Device Problem		
					CT981	Single use	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem		
							F1901	Additional Surgery	A24	No Apparent Device Problem		
					CT983	Surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem		
							F1901	Additional Surgery	A24	No Apparent Device Problem		
12691	2023		47265	Dual-chamber implantable pacemaker, rate-responsive	CT141	Electrodes/Leads and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem		
							F1901	Additional Surgery	A24	No Apparent Device Problem		
					CT145	Cardiovascular devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem		
							F1901	Additional Surgery	A24	No Apparent Device Problem		
					CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem		
							F1901	Additional Surgery	A24	No Apparent Device Problem		
					CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem		
							F1901	Additional Surgery	A24	No Apparent Device Problem		
					CT233	Surgical	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem		
							F1901	Additional Surgery	A24	No Apparent Device Problem		
					CT2406	Implantable	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem		
							F1901	Additional Surgery	A24	No Apparent Device Problem		
					CT2407	Active implantable	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem		
							F1901	Additional Surgery	A24	No Apparent Device Problem		
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem		
							F1901	Additional Surgery	A24	No Apparent Device Problem		
					CT2727	Active implant devices/ancillaries	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem		
							F1901	Additional Surgery	A24	No Apparent Device Problem		
					CT2750	Implantable cardiac electrical stimulators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem		
							F1901	Additional Surgery	A24	No Apparent Device Problem		
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem		
							F1901	Additional Surgery	A24	No Apparent Device Problem		
					CT2985	Active	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem		
							F1901	Additional Surgery	A24	No Apparent Device Problem		
					CT3001	Implantable metal	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem		
							F1901	Additional Surgery	A24	No Apparent Device Problem		
					CT3002	Implantable synthetic polymer	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem		
							F1901	Additional Surgery	A24	No Apparent Device Problem		
					CT3006	Implantable inorganic material	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem		
							F1901	Additional Surgery	A24	No Apparent Device Problem		
					CT314	Electrically powered	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem		
							F1901	Additional Surgery	A24	No Apparent Device Problem		
					CT321	Long-term surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem		
							F1901	Additional Surgery	A24	No Apparent Device Problem		
					CT449	Generators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem		
							F1901	Additional Surgery	A24	No Apparent Device Problem		
					CT981	Single use	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem		
							F1901	Additional Surgery	A24	No Apparent Device Problem		
					CT983	Surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem		
							F1901	Additional Surgery	A24	No Apparent Device Problem		
12692	2023		47265	Dual-chamber implantable pacemaker, rate-responsive	CT141	Electrodes/Leads and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem		
							F1901	Additional Surgery	A24	No Apparent Device Problem		
					CT145	Cardiovascular devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem		
							F1901	Additional Surgery	A24	No Apparent Device Problem		
					CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem		
							F1901	Additional Surgery	A24	No Apparent Device Problem		
					CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem		
							F1901	Additional Surgery	A24	No Apparent Device Problem		
					CT233	Surgical	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem		
							F1901	Additional Surgery	A24	No Apparent Device Problem		
					CT2406	Implantable	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem		
							F1901	Additional Surgery	A24	No Apparent Device Problem		
					CT2407	Active implantable	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem		
							F1901	Additional Surgery	A24	No Apparent Device Problem		
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem		
							F1901	Additional Surgery	A24	No Apparent Device Problem		
					CT2727	Active implant devices/ancillaries	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem		
							F1901	Additional Surgery	A24	No Apparent Device Problem		
					CT2750	Implantable cardiac electrical stimulators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem		

							F1901	Additional Surgery	A24	No Apparent Device Problem		
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem		
							F1901	Additional Surgery	A24	No Apparent Device Problem		
					CT2985	Active	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem		
							F1901	Additional Surgery	A24	No Apparent Device Problem		
					CT3001	Implantable metal	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem		
							F1901	Additional Surgery	A24	No Apparent Device Problem		
					CT3002	Implantable synthetic polymer	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem		
							F1901	Additional Surgery	A24	No Apparent Device Problem		
					CT3006	Implantable inorganic material	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem		
							F1901	Additional Surgery	A24	No Apparent Device Problem		
					CT314	Electrically powered	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem		
							F1901	Additional Surgery	A24	No Apparent Device Problem		
					CT321	Long-term surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem		
							F1901	Additional Surgery	A24	No Apparent Device Problem		
					CT449	Generators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem		
							F1901	Additional Surgery	A24	No Apparent Device Problem		
					CT981	Single use	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem		
							F1901	Additional Surgery	A24	No Apparent Device Problem		
					CT983	Surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem		
							F1901	Additional Surgery	A24	No Apparent Device Problem		
12693	2023		47265	Dual-chamber implantable pacemaker, rate-responsive	CT141	Electrodes/Leads and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem		
							F1901	Additional Surgery	A24	No Apparent Device Problem		
					CT145	Cardiovascular devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem		
							F1901	Additional Surgery	A24	No Apparent Device Problem		
					CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem		
							F1901	Additional Surgery	A24	No Apparent Device Problem		
					CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem		
							F1901	Additional Surgery	A24	No Apparent Device Problem		
					CT233	Surgical	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem		
							F1901	Additional Surgery	A24	No Apparent Device Problem		
					CT2406	Implantable	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem		
							F1901	Additional Surgery	A24	No Apparent Device Problem		
					CT2407	Active implantable	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem		
							F1901	Additional Surgery	A24	No Apparent Device Problem		
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem		
							F1901	Additional Surgery	A24	No Apparent Device Problem		
					CT2727	Active implant devices/ancillaries	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem		
							F1901	Additional Surgery	A24	No Apparent Device Problem		
					CT2750	Implantable cardiac electrical stimulators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem		
							F1901	Additional Surgery	A24	No Apparent Device Problem		
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem		
							F1901	Additional Surgery	A24	No Apparent Device Problem		
					CT2985	Active	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem		
							F1901	Additional Surgery	A24	No Apparent Device Problem		
					CT3001	Implantable metal	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem		
							F1901	Additional Surgery	A24	No Apparent Device Problem		
					CT3002	Implantable synthetic polymer	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem		
							F1901	Additional Surgery	A24	No Apparent Device Problem		
					CT3006	Implantable inorganic material	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem		
							F1901	Additional Surgery	A24	No Apparent Device Problem		
					CT314	Electrically powered	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem		
							F1901	Additional Surgery	A24	No Apparent Device Problem		
					CT321	Long-term surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem		
							F1901	Additional Surgery	A24	No Apparent Device Problem		
					CT449	Generators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem		
							F1901	Additional Surgery	A24	No Apparent Device Problem		
					CT981	Single use	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem		
							F1901	Additional Surgery	A24	No Apparent Device Problem		
					CT983	Surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem		
							F1901	Additional Surgery	A24	No Apparent Device Problem		
12694	2023		47265	Dual-chamber implantable pacemaker, rate-responsive	CT141	Electrodes/Leads and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem		
							F1901	Additional Surgery	A24	No Apparent Device Problem		
					CT145	Cardiovascular devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem		
							F1901	Additional Surgery	A24	No Apparent Device Problem		
					CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem		
							F1901	Additional Surgery	A24	No Apparent Device Problem		
					CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem		
							F1901	Additional Surgery	A24	No Apparent Device Problem		
					CT233	Surgical	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem		
							F1901	Additional Surgery	A24	No Apparent Device Problem		
					CT2406	Implantable	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem		
							F1901	Additional Surgery	A24	No Apparent Device Problem		
					CT2407	Active implantable	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem		
							F1901	Additional Surgery	A24	No Apparent Device Problem		
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem		
							F1901	Additional Surgery	A24	No Apparent Device Problem		
					CT2727	Active implant devices/ancillaries	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem		
							F1901	Additional Surgery	A24	No Apparent Device Problem		
					CT2750	Implantable cardiac electrical stimulators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem		
							F1901	Additional Surgery	A24	No Apparent Device Problem		
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem		
							F1901	Additional Surgery	A24	No Apparent Device Problem		
					CT2985	Active	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem		
							F1901	Additional Surgery	A24	No Apparent Device Problem		
					CT3001	Implantable metal	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem		
							F1901	Additional Surgery	A24	No Apparent Device Problem		
					CT3002	Implantable synthetic polymer	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem		

						F1901	Additional Surgery	A24	No Apparent Device Problem			
				CT3006	Implantable inorganic material	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem			
						F1901	Additional Surgery	A24	No Apparent Device Problem			
				CT314	Electrically powered	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem			
						F1901	Additional Surgery	A24	No Apparent Device Problem			
				CT321	Long-term surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem			
						F1901	Additional Surgery	A24	No Apparent Device Problem			
				CT449	Generators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem			
						F1901	Additional Surgery	A24	No Apparent Device Problem			
				CT981	Single use	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem			
						F1901	Additional Surgery	A24	No Apparent Device Problem			
				CT983	Surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem			
						F1901	Additional Surgery	A24	No Apparent Device Problem			
12695	2023		47265	Dual-chamber implantable pacemaker, rate-responsive	CT141	Electrodes/Leads and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem		
							F1901	Additional Surgery	A24	No Apparent Device Problem		
					CT145	Cardiovascular devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem		
							F1901	Additional Surgery	A24	No Apparent Device Problem		
					CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem		
							F1901	Additional Surgery	A24	No Apparent Device Problem		
					CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem		
							F1901	Additional Surgery	A24	No Apparent Device Problem		
					CT233	Surgical	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem		
							F1901	Additional Surgery	A24	No Apparent Device Problem		
					CT2406	Implantable	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem		
							F1901	Additional Surgery	A24	No Apparent Device Problem		
					CT2407	Active implantable	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem		
							F1901	Additional Surgery	A24	No Apparent Device Problem		
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem		
							F1901	Additional Surgery	A24	No Apparent Device Problem		
					CT2727	Active implant devices/ancillaries	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem		
							F1901	Additional Surgery	A24	No Apparent Device Problem		
					CT2750	Implantable cardiac electrical stimulators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem		
							F1901	Additional Surgery	A24	No Apparent Device Problem		
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem		
							F1901	Additional Surgery	A24	No Apparent Device Problem		
					CT2985	Active	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem		
							F1901	Additional Surgery	A24	No Apparent Device Problem		
					CT3001	Implantable metal	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem		
							F1901	Additional Surgery	A24	No Apparent Device Problem		
					CT3002	Implantable synthetic polymer	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem		
							F1901	Additional Surgery	A24	No Apparent Device Problem		
					CT3006	Implantable inorganic material	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem		
							F1901	Additional Surgery	A24	No Apparent Device Problem		
					CT314	Electrically powered	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem		
							F1901	Additional Surgery	A24	No Apparent Device Problem		
					CT321	Long-term surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem		
							F1901	Additional Surgery	A24	No Apparent Device Problem		
					CT449	Generators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem		
							F1901	Additional Surgery	A24	No Apparent Device Problem		
					CT981	Single use	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem		
							F1901	Additional Surgery	A24	No Apparent Device Problem		
					CT983	Surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem		
							F1901	Additional Surgery	A24	No Apparent Device Problem		
12696	2023		47265	Dual-chamber implantable pacemaker, rate-responsive	CT141	Electrodes/Leads and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem		
							F1901	Additional Surgery	A24	No Apparent Device Problem		
					CT145	Cardiovascular devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem		
							F1901	Additional Surgery	A24	No Apparent Device Problem		
					CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem		
							F1901	Additional Surgery	A24	No Apparent Device Problem		
					CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem		
							F1901	Additional Surgery	A24	No Apparent Device Problem		
					CT233	Surgical	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem		
							F1901	Additional Surgery	A24	No Apparent Device Problem		
					CT2406	Implantable	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem		
							F1901	Additional Surgery	A24	No Apparent Device Problem		
					CT2407	Active implantable	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem		
							F1901	Additional Surgery	A24	No Apparent Device Problem		
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem		
							F1901	Additional Surgery	A24	No Apparent Device Problem		
					CT2727	Active implant devices/ancillaries	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem		
							F1901	Additional Surgery	A24	No Apparent Device Problem		
					CT2750	Implantable cardiac electrical stimulators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem		
							F1901	Additional Surgery	A24	No Apparent Device Problem		
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem		
							F1901	Additional Surgery	A24	No Apparent Device Problem		
					CT2985	Active	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem		
							F1901	Additional Surgery	A24	No Apparent Device Problem		
					CT3001	Implantable metal	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem		
							F1901	Additional Surgery	A24	No Apparent Device Problem		
					CT3002	Implantable synthetic polymer	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem		
							F1901	Additional Surgery	A24	No Apparent Device Problem		
					CT3006	Implantable inorganic material	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem		
							F1901	Additional Surgery	A24	No Apparent Device Problem		
					CT314	Electrically powered	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem		
							F1901	Additional Surgery	A24	No Apparent Device Problem		
					CT321	Long-term surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem		
							F1901	Additional Surgery	A24	No Apparent Device Problem		
					CT449	Generators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem		

							F1901	Additional Surgery	A24	No Apparent Device Problem			
					CT981	Single use	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem			
							F1901	Additional Surgery	A24	No Apparent Device Problem			
					CT983	Surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem			
							F1901	Additional Surgery	A24	No Apparent Device Problem			
12697	2023		47265	Dual-chamber implantable pacemaker, rate-responsive	CT141	Electrodes/Leads and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem			
							F1901	Additional Surgery	A24	No Apparent Device Problem			
					CT145	Cardiovascular devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem			
							F1901	Additional Surgery	A24	No Apparent Device Problem			
					CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem			
							F1901	Additional Surgery	A24	No Apparent Device Problem			
					CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem			
							F1901	Additional Surgery	A24	No Apparent Device Problem			
					CT233	Surgical	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem			
							F1901	Additional Surgery	A24	No Apparent Device Problem			
					CT2406	Implantable	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem			
							F1901	Additional Surgery	A24	No Apparent Device Problem			
					CT2407	Active implantable	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem			
							F1901	Additional Surgery	A24	No Apparent Device Problem			
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem			
							F1901	Additional Surgery	A24	No Apparent Device Problem			
					CT2727	Active implant devices/ancillaries	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem			
							F1901	Additional Surgery	A24	No Apparent Device Problem			
					CT2750	Implantable cardiac electrical stimulators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem			
							F1901	Additional Surgery	A24	No Apparent Device Problem			
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem			
							F1901	Additional Surgery	A24	No Apparent Device Problem			
					CT2985	Active	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem			
							F1901	Additional Surgery	A24	No Apparent Device Problem			
					CT3001	Implantable metal	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem			
							F1901	Additional Surgery	A24	No Apparent Device Problem			
					CT3002	Implantable synthetic polymer	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem			
							F1901	Additional Surgery	A24	No Apparent Device Problem			
					CT3006	Implantable inorganic material	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem			
							F1901	Additional Surgery	A24	No Apparent Device Problem			
					CT314	Electrically powered	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem			
							F1901	Additional Surgery	A24	No Apparent Device Problem			
					CT321	Long-term surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem			
							F1901	Additional Surgery	A24	No Apparent Device Problem			
					CT449	Generators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem			
							F1901	Additional Surgery	A24	No Apparent Device Problem			
					CT981	Single use	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem			
							F1901	Additional Surgery	A24	No Apparent Device Problem			
					CT983	Surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem			
							F1901	Additional Surgery	A24	No Apparent Device Problem			
12698	2023	All other reportable incidents	47265	Dual-chamber implantable pacemaker, rate-responsive	CT141	Electrodes/Leads and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	06/03/2023	06/03/2023	
							F1901	Additional Surgery	A24	No Apparent Device Problem	06/03/2023	06/03/2023	
					CT145	Cardiovascular devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	06/03/2023	06/03/2023	
							F1901	Additional Surgery	A24	No Apparent Device Problem	06/03/2023	06/03/2023	
					CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	06/03/2023	06/03/2023	
							F1901	Additional Surgery	A24	No Apparent Device Problem	06/03/2023	06/03/2023	
					CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	06/03/2023	06/03/2023	
							F1901	Additional Surgery	A24	No Apparent Device Problem	06/03/2023	06/03/2023	
					CT233	Surgical	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	06/03/2023	06/03/2023	
							F1901	Additional Surgery	A24	No Apparent Device Problem	06/03/2023	06/03/2023	
					CT2406	Implantable	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	06/03/2023	06/03/2023	
							F1901	Additional Surgery	A24	No Apparent Device Problem	06/03/2023	06/03/2023	
					CT2407	Active implantable	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	06/03/2023	06/03/2023	
							F1901	Additional Surgery	A24	No Apparent Device Problem	06/03/2023	06/03/2023	
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	06/03/2023	06/03/2023	
							F1901	Additional Surgery	A24	No Apparent Device Problem	06/03/2023	06/03/2023	
					CT2727	Active implant devices/ancillaries	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	06/03/2023	06/03/2023	
							F1901	Additional Surgery	A24	No Apparent Device Problem	06/03/2023	06/03/2023	
					CT2750	Implantable cardiac electrical stimulators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	06/03/2023	06/03/2023	
							F1901	Additional Surgery	A24	No Apparent Device Problem	06/03/2023	06/03/2023	
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	06/03/2023	06/03/2023	
							F1901	Additional Surgery	A24	No Apparent Device Problem	06/03/2023	06/03/2023	
					CT2985	Active	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	06/03/2023	06/03/2023	
							F1901	Additional Surgery	A24	No Apparent Device Problem	06/03/2023	06/03/2023	
					CT3001	Implantable metal	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	06/03/2023	06/03/2023	
							F1901	Additional Surgery	A24	No Apparent Device Problem	06/03/2023	06/03/2023	
					CT3002	Implantable synthetic polymer	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	06/03/2023	06/03/2023	
							F1901	Additional Surgery	A24	No Apparent Device Problem	06/03/2023	06/03/2023	
					CT3006	Implantable inorganic material	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	06/03/2023	06/03/2023	
							F1901	Additional Surgery	A24	No Apparent Device Problem	06/03/2023	06/03/2023	
					CT314	Electrically powered	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	06/03/2023	06/03/2023	
							F1901	Additional Surgery	A24	No Apparent Device Problem	06/03/2023	06/03/2023	
					CT321	Long-term surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	06/03/2023	06/03/2023	
							F1901	Additional Surgery	A24	No Apparent Device Problem	06/03/2023	06/03/2023	
					CT449	Generators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	06/03/2023	06/03/2023	
							F1901	Additional Surgery	A24	No Apparent Device Problem	06/03/2023	06/03/2023	
					CT981	Single use	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	06/03/2023	06/03/2023	
							F1901	Additional Surgery	A24	No Apparent Device Problem	06/03/2023	06/03/2023	
					CT983	Surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	06/03/2023	06/03/2023	
							F1901	Additional Surgery	A24	No Apparent Device Problem	06/03/2023	06/03/2023	
12699	2023	All other reportable incidents	35223	Endocardial/interventricular septal pacing lead	CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070909	Oversensing			
							F1901	Additional Surgery	A070909	Oversensing			
12700	2023	All other reportable incidents	35852	Single-chamber implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery			

									A0908	Incorrect, Inadequate or Imprecise Result or Readings		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
									A0908	Incorrect, Inadequate or Imprecise Result or Readings		
12701	2023	All other reportable incidents	47265	Dual-chamber implantable pacemaker, rate-responsive	CT141	Electrodes/Leads and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT145	Cardiovascular devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT233	Surgical	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT2406	Implantable	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT2407	Active implantable	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT2727	Active implant devices/ancillaries	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT2750	Implantable cardiac electrical stimulators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT2985	Active	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT3001	Implantable metal	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT3002	Implantable synthetic polymer	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT3006	Implantable inorganic material	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT314	Electrically powered	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT321	Long-term surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT449	Generators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT981	Single use	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT983	Surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
12702	2023	All other reportable incidents	37265	Dual-chamber implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E2104	Electric Shock	A071302	Inappropriate/Inadequate Shock/Stimulation		
							F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A071302	Inappropriate/Inadequate Shock/Stimulation		
12703	2023	All other reportable incidents	35852	Single-chamber implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
12704	2023	All other reportable incidents	35852	Single-chamber implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E2330	Pain	A010402	Migration	25/01/2022	25/01/2022
							F08	Hospitalization or Prolonged Hospitalization	A010402	Migration	25/01/2022	25/01/2022
							F1905	Device Revision or Replacement	A010402	Migration	25/01/2022	25/01/2022
12705	2023	All other reportable incidents	47265	Dual-chamber implantable pacemaker, rate-responsive	CT141	Electrodes/Leads and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070910	Undersensing		
							F15	Recognized Device or Procedural Complication	A070910	Undersensing		
					CT145	Cardiovascular devices	E2403	No Clinical Signs, Symptoms or Conditions	A070910	Undersensing		
							F15	Recognized Device or Procedural Complication	A070910	Undersensing		
					CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070910	Undersensing		
							F15	Recognized Device or Procedural Complication	A070910	Undersensing		
					CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A070910	Undersensing		
							F15	Recognized Device or Procedural Complication	A070910	Undersensing		
					CT233	Surgical	E2403	No Clinical Signs, Symptoms or Conditions	A070910	Undersensing		
							F15	Recognized Device or Procedural Complication	A070910	Undersensing		
					CT2406	Implantable	E2403	No Clinical Signs, Symptoms or Conditions	A070910	Undersensing		
							F15	Recognized Device or Procedural Complication	A070910	Undersensing		
					CT2407	Active implantable	E2403	No Clinical Signs, Symptoms or Conditions	A070910	Undersensing		
							F15	Recognized Device or Procedural Complication	A070910	Undersensing		
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A070910	Undersensing		
							F15	Recognized Device or Procedural Complication	A070910	Undersensing		
					CT2727	Active implant devices/ancillaries	E2403	No Clinical Signs, Symptoms or Conditions	A070910	Undersensing		
							F15	Recognized Device or Procedural Complication	A070910	Undersensing		
					CT2750	Implantable cardiac electrical stimulators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070910	Undersensing		
							F15	Recognized Device or Procedural Complication	A070910	Undersensing		
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A070910	Undersensing		
							F15	Recognized Device or Procedural Complication	A070910	Undersensing		
					CT2985	Active	E2403	No Clinical Signs, Symptoms or Conditions	A070910	Undersensing		
							F15	Recognized Device or Procedural Complication	A070910	Undersensing		
					CT3001	Implantable metal	E2403	No Clinical Signs, Symptoms or Conditions	A070910	Undersensing		
							F15	Recognized Device or Procedural Complication	A070910	Undersensing		
					CT3002	Implantable synthetic polymer	E2403	No Clinical Signs, Symptoms or Conditions	A070910	Undersensing		
							F15	Recognized Device or Procedural Complication	A070910	Undersensing		
					CT3006	Implantable inorganic material	E2403	No Clinical Signs, Symptoms or Conditions	A070910	Undersensing		
							F15	Recognized Device or Procedural Complication	A070910	Undersensing		
					CT314	Electrically powered	E2403	No Clinical Signs, Symptoms or Conditions	A070910	Undersensing		
							F15	Recognized Device or Procedural Complication	A070910	Undersensing		
					CT321	Long-term surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A070910	Undersensing		
							F15	Recognized Device or Procedural Complication	A070910	Undersensing		

					CT449	Generators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070910	Undersensing		
							F15	Recognized Device or Procedural Complication	A070910	Undersensing		
					CT981	Single use	E2403	No Clinical Signs, Symptoms or Conditions	A070910	Undersensing		
							F15	Recognized Device or Procedural Complication	A070910	Undersensing		
					CT983	Surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A070910	Undersensing		
							F15	Recognized Device or Procedural Complication	A070910	Undersensing		
12706	2023	All other reportable incidents	47265	Dual-chamber implantable pacemaker, rate-responsive	CT141	Electrodes/Leads and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	13/02/2023	13/02/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	13/02/2023	13/02/2023
					CT145	Cardiovascular devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	13/02/2023	13/02/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	13/02/2023	13/02/2023
					CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	13/02/2023	13/02/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	13/02/2023	13/02/2023
					CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	13/02/2023	13/02/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	13/02/2023	13/02/2023
					CT233	Surgical	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	13/02/2023	13/02/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	13/02/2023	13/02/2023
					CT2406	Implantable	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	13/02/2023	13/02/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	13/02/2023	13/02/2023
					CT2407	Active implantable	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	13/02/2023	13/02/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	13/02/2023	13/02/2023
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	13/02/2023	13/02/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	13/02/2023	13/02/2023
					CT2727	Active implant devices/ancillaries	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	13/02/2023	13/02/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	13/02/2023	13/02/2023
					CT2750	Implantable cardiac electrical stimulators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	13/02/2023	13/02/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	13/02/2023	13/02/2023
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	13/02/2023	13/02/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	13/02/2023	13/02/2023
					CT2985	Active	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	13/02/2023	13/02/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	13/02/2023	13/02/2023
					CT3001	Implantable metal	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	13/02/2023	13/02/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	13/02/2023	13/02/2023
					CT3002	Implantable synthetic polymer	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	13/02/2023	13/02/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	13/02/2023	13/02/2023
					CT3006	Implantable inorganic material	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	13/02/2023	13/02/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	13/02/2023	13/02/2023
					CT314	Electrically powered	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	13/02/2023	13/02/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	13/02/2023	13/02/2023
					CT321	Long-term surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	13/02/2023	13/02/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	13/02/2023	13/02/2023
					CT449	Generators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	13/02/2023	13/02/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	13/02/2023	13/02/2023
					CT981	Single use	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	13/02/2023	13/02/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	13/02/2023	13/02/2023
					CT983	Surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	13/02/2023	13/02/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	13/02/2023	13/02/2023
12707	2023		47265	Dual-chamber implantable pacemaker, rate-responsive	CT141	Electrodes/Leads and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	03/03/2023	03/03/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	03/03/2023	03/03/2023
					CT145	Cardiovascular devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	03/03/2023	03/03/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	03/03/2023	03/03/2023
					CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	03/03/2023	03/03/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	03/03/2023	03/03/2023
					CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	03/03/2023	03/03/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	03/03/2023	03/03/2023
					CT233	Surgical	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	03/03/2023	03/03/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	03/03/2023	03/03/2023
					CT2406	Implantable	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	03/03/2023	03/03/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	03/03/2023	03/03/2023
					CT2407	Active implantable	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	03/03/2023	03/03/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	03/03/2023	03/03/2023
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	03/03/2023	03/03/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	03/03/2023	03/03/2023
					CT2727	Active implant devices/ancillaries	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	03/03/2023	03/03/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	03/03/2023	03/03/2023
					CT2750	Implantable cardiac electrical stimulators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	03/03/2023	03/03/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	03/03/2023	03/03/2023
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	03/03/2023	03/03/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	03/03/2023	03/03/2023
					CT2985	Active	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	03/03/2023	03/03/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	03/03/2023	03/03/2023
					CT3001	Implantable metal	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	03/03/2023	03/03/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	03/03/2023	03/03/2023
					CT3002	Implantable synthetic polymer	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	03/03/2023	03/03/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	03/03/2023	03/03/2023
					CT3006	Implantable inorganic material	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	03/03/2023	03/03/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	03/03/2023	03/03/2023
					CT314	Electrically powered	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	03/03/2023	03/03/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	03/03/2023	03/03/2023
					CT321	Long-term surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	03/03/2023	03/03/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	03/03/2023	03/03/2023
					CT449	Generators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	03/03/2023	03/03/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	03/03/2023	03/03/2023
					CT981	Single use	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	03/03/2023	03/03/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	03/03/2023	03/03/2023
					CT983	Surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	03/03/2023	03/03/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	03/03/2023	03/03/2023
12708	2023		47265	Dual-chamber implantable pacemaker, rate-responsive	CT141	Electrodes/Leads and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem		
							F1901	Additional Surgery	A24	No Apparent Device Problem		

					CT145	Cardiovascular devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem		
							F1901	Additional Surgery	A24	No Apparent Device Problem		
					CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem		
							F1901	Additional Surgery	A24	No Apparent Device Problem		
					CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem		
							F1901	Additional Surgery	A24	No Apparent Device Problem		
					CT233	Surgical	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem		
							F1901	Additional Surgery	A24	No Apparent Device Problem		
					CT2406	Implantable	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem		
							F1901	Additional Surgery	A24	No Apparent Device Problem		
					CT2407	Active implantable	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem		
							F1901	Additional Surgery	A24	No Apparent Device Problem		
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem		
							F1901	Additional Surgery	A24	No Apparent Device Problem		
					CT2727	Active implant devices/ancillaries	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem		
							F1901	Additional Surgery	A24	No Apparent Device Problem		
					CT2750	Implantable cardiac electrical stimulators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem		
							F1901	Additional Surgery	A24	No Apparent Device Problem		
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem		
							F1901	Additional Surgery	A24	No Apparent Device Problem		
					CT2985	Active	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem		
							F1901	Additional Surgery	A24	No Apparent Device Problem		
					CT3001	Implantable metal	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem		
							F1901	Additional Surgery	A24	No Apparent Device Problem		
					CT3002	Implantable synthetic polymer	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem		
							F1901	Additional Surgery	A24	No Apparent Device Problem		
					CT3006	Implantable inorganic material	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem		
							F1901	Additional Surgery	A24	No Apparent Device Problem		
					CT314	Electrically powered	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem		
							F1901	Additional Surgery	A24	No Apparent Device Problem		
					CT321	Long-term surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem		
							F1901	Additional Surgery	A24	No Apparent Device Problem		
					CT449	Generators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem		
							F1901	Additional Surgery	A24	No Apparent Device Problem		
					CT981	Single use	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem		
							F1901	Additional Surgery	A24	No Apparent Device Problem		
					CT983	Surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem		
							F1901	Additional Surgery	A24	No Apparent Device Problem		
12709	2023	All other reportable incidents	35223	Endocardial/interventricular septal pacing lead	CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070909	Oversensing		
							F26	No Health Consequences or Impact	A070909	Oversensing		
12710	2023	All other reportable incidents	35852	Single-chamber implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
									A0908	Incorrect, Inadequate or Imprecise Result or Readings		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
									A0908	Incorrect, Inadequate or Imprecise Result or Readings		
12711	2023		47265	Dual-chamber implantable pacemaker, rate-responsive	CT141	Electrodes/Leads and associated devices	E2401	Insufficient Information	A070504	Premature Discharge of Battery	20/03/2023	
							F1903	Device Explantation	A070504	Premature Discharge of Battery	20/03/2023	
					CT145	Cardiovascular devices	E2401	Insufficient Information	A070504	Premature Discharge of Battery	20/03/2023	
							F1903	Device Explantation	A070504	Premature Discharge of Battery	20/03/2023	
					CT1684	Cardiac pacemakers and associated devices	E2401	Insufficient Information	A070504	Premature Discharge of Battery	20/03/2023	
							F1903	Device Explantation	A070504	Premature Discharge of Battery	20/03/2023	
					CT1685	Cardiac pacemakers	E2401	Insufficient Information	A070504	Premature Discharge of Battery	20/03/2023	
							F1903	Device Explantation	A070504	Premature Discharge of Battery	20/03/2023	
					CT233	Surgical	E2401	Insufficient Information	A070504	Premature Discharge of Battery	20/03/2023	
							F1903	Device Explantation	A070504	Premature Discharge of Battery	20/03/2023	
					CT2406	Implantable	E2401	Insufficient Information	A070504	Premature Discharge of Battery	20/03/2023	
							F1903	Device Explantation	A070504	Premature Discharge of Battery	20/03/2023	
					CT2407	Active implantable	E2401	Insufficient Information	A070504	Premature Discharge of Battery	20/03/2023	
							F1903	Device Explantation	A070504	Premature Discharge of Battery	20/03/2023	
					CT257	Implantable cardiac pacemakers	E2401	Insufficient Information	A070504	Premature Discharge of Battery	20/03/2023	
							F1903	Device Explantation	A070504	Premature Discharge of Battery	20/03/2023	
					CT2727	Active implant devices/ancillaries	E2401	Insufficient Information	A070504	Premature Discharge of Battery	20/03/2023	
							F1903	Device Explantation	A070504	Premature Discharge of Battery	20/03/2023	
					CT2750	Implantable cardiac electrical stimulators and associated devices	E2401	Insufficient Information	A070504	Premature Discharge of Battery	20/03/2023	
							F1903	Device Explantation	A070504	Premature Discharge of Battery	20/03/2023	
					CT2866	Implantable cardiac electrical stimulators	E2401	Insufficient Information	A070504	Premature Discharge of Battery	20/03/2023	
							F1903	Device Explantation	A070504	Premature Discharge of Battery	20/03/2023	
					CT2985	Active	E2401	Insufficient Information	A070504	Premature Discharge of Battery	20/03/2023	
							F1903	Device Explantation	A070504	Premature Discharge of Battery	20/03/2023	
					CT3001	Implantable metal	E2401	Insufficient Information	A070504	Premature Discharge of Battery	20/03/2023	
							F1903	Device Explantation	A070504	Premature Discharge of Battery	20/03/2023	
					CT3002	Implantable synthetic polymer	E2401	Insufficient Information	A070504	Premature Discharge of Battery	20/03/2023	
							F1903	Device Explantation	A070504	Premature Discharge of Battery	20/03/2023	
					CT3006	Implantable inorganic material	E2401	Insufficient Information	A070504	Premature Discharge of Battery	20/03/2023	
							F1903	Device Explantation	A070504	Premature Discharge of Battery	20/03/2023	
					CT314	Electrically powered	E2401	Insufficient Information	A070504	Premature Discharge of Battery	20/03/2023	
							F1903	Device Explantation	A070504	Premature Discharge of Battery	20/03/2023	
					CT321	Long-term surgical invasive	E2401	Insufficient Information	A070504	Premature Discharge of Battery	20/03/2023	
							F1903	Device Explantation	A070504	Premature Discharge of Battery	20/03/2023	
					CT449	Generators and associated devices	E2401	Insufficient Information	A070504	Premature Discharge of Battery	20/03/2023	
							F1903	Device Explantation	A070504	Premature Discharge of Battery	20/03/2023	
					CT981	Single use	E2401	Insufficient Information	A070504	Premature Discharge of Battery	20/03/2023	
							F1903	Device Explantation	A070504	Premature Discharge of Battery	20/03/2023	
					CT983	Surgical invasive	E2401	Insufficient Information	A070504	Premature Discharge of Battery	20/03/2023	
							F1903	Device Explantation	A070504	Premature Discharge of Battery	20/03/2023	
12712	2023	All other reportable incidents	35223	Endocardial/interventricular septal pacing lead	CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070909	Oversensing		
							F1901	Additional Surgery	A070909	Oversensing		

12713	2023		47265	Dual-chamber implantable pacemaker, rate-responsive	CT141	Electrodes/Leads and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	16/03/2023	16/03/2023
							F1908	Prolonged Surgery	A24	No Apparent Device Problem	16/03/2023	16/03/2023
					CT145	Cardiovascular devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	16/03/2023	16/03/2023
							F1908	Prolonged Surgery	A24	No Apparent Device Problem	16/03/2023	16/03/2023
					CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	16/03/2023	16/03/2023
							F1908	Prolonged Surgery	A24	No Apparent Device Problem	16/03/2023	16/03/2023
					CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	16/03/2023	16/03/2023
							F1908	Prolonged Surgery	A24	No Apparent Device Problem	16/03/2023	16/03/2023
					CT233	Surgical	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	16/03/2023	16/03/2023
							F1908	Prolonged Surgery	A24	No Apparent Device Problem	16/03/2023	16/03/2023
					CT2406	Implantable	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	16/03/2023	16/03/2023
							F1908	Prolonged Surgery	A24	No Apparent Device Problem	16/03/2023	16/03/2023
					CT2407	Active implantable	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	16/03/2023	16/03/2023
							F1908	Prolonged Surgery	A24	No Apparent Device Problem	16/03/2023	16/03/2023
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	16/03/2023	16/03/2023
							F1908	Prolonged Surgery	A24	No Apparent Device Problem	16/03/2023	16/03/2023
					CT2727	Active implant devices/ancillaries	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	16/03/2023	16/03/2023
							F1908	Prolonged Surgery	A24	No Apparent Device Problem	16/03/2023	16/03/2023
					CT2750	Implantable cardiac electrical stimulators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	16/03/2023	16/03/2023
							F1908	Prolonged Surgery	A24	No Apparent Device Problem	16/03/2023	16/03/2023
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	16/03/2023	16/03/2023
							F1908	Prolonged Surgery	A24	No Apparent Device Problem	16/03/2023	16/03/2023
					CT2985	Active	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	16/03/2023	16/03/2023
							F1908	Prolonged Surgery	A24	No Apparent Device Problem	16/03/2023	16/03/2023
					CT3001	Implantable metal	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	16/03/2023	16/03/2023
							F1908	Prolonged Surgery	A24	No Apparent Device Problem	16/03/2023	16/03/2023
					CT3002	Implantable synthetic polymer	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	16/03/2023	16/03/2023
							F1908	Prolonged Surgery	A24	No Apparent Device Problem	16/03/2023	16/03/2023
					CT3006	Implantable inorganic material	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	16/03/2023	16/03/2023
							F1908	Prolonged Surgery	A24	No Apparent Device Problem	16/03/2023	16/03/2023
					CT314	Electrically powered	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	16/03/2023	16/03/2023
							F1908	Prolonged Surgery	A24	No Apparent Device Problem	16/03/2023	16/03/2023
					CT321	Long-term surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	16/03/2023	16/03/2023
							F1908	Prolonged Surgery	A24	No Apparent Device Problem	16/03/2023	16/03/2023
					CT449	Generators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	16/03/2023	16/03/2023
							F1908	Prolonged Surgery	A24	No Apparent Device Problem	16/03/2023	16/03/2023
					CT981	Single use	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	16/03/2023	16/03/2023
							F1908	Prolonged Surgery	A24	No Apparent Device Problem	16/03/2023	16/03/2023
					CT983	Surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	16/03/2023	16/03/2023
							F1908	Prolonged Surgery	A24	No Apparent Device Problem	16/03/2023	16/03/2023
12714	2023	All other reportable incidents	17882	Physiologic-monitoring defibrillation system	CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070908	Failure to Sense		
							F26	No Health Consequences or Impact	A070908	Failure to Sense		
12715	2023	All other reportable incidents	47265	Dual-chamber implantable pacemaker, rate-responsive	CT141	Electrodes/Leads and associated devices	E0112	Dizziness	A071102	Failure to Interrogate	28/03/2023	28/03/2023
									A071204	No Pacing	28/03/2023	28/03/2023
							E0601	Arrhythmia	A071102	Failure to Interrogate	28/03/2023	28/03/2023
									A071204	No Pacing	28/03/2023	28/03/2023
							F1901	Additional Surgery	A071102	Failure to Interrogate	28/03/2023	28/03/2023
									A071204	No Pacing	28/03/2023	28/03/2023
					CT145	Cardiovascular devices	E0112	Dizziness	A071102	Failure to Interrogate	28/03/2023	28/03/2023
									A071204	No Pacing	28/03/2023	28/03/2023
							E0601	Arrhythmia	A071102	Failure to Interrogate	28/03/2023	28/03/2023
									A071204	No Pacing	28/03/2023	28/03/2023
							F1901	Additional Surgery	A071102	Failure to Interrogate	28/03/2023	28/03/2023
									A071204	No Pacing	28/03/2023	28/03/2023
					CT1684	Cardiac pacemakers and associated devices	E0112	Dizziness	A071102	Failure to Interrogate	28/03/2023	28/03/2023
									A071204	No Pacing	28/03/2023	28/03/2023
							E0601	Arrhythmia	A071102	Failure to Interrogate	28/03/2023	28/03/2023
									A071204	No Pacing	28/03/2023	28/03/2023
							F1901	Additional Surgery	A071102	Failure to Interrogate	28/03/2023	28/03/2023
									A071204	No Pacing	28/03/2023	28/03/2023
					CT1685	Cardiac pacemakers	E0112	Dizziness	A071102	Failure to Interrogate	28/03/2023	28/03/2023
									A071204	No Pacing	28/03/2023	28/03/2023
							E0601	Arrhythmia	A071102	Failure to Interrogate	28/03/2023	28/03/2023
									A071204	No Pacing	28/03/2023	28/03/2023
							F1901	Additional Surgery	A071102	Failure to Interrogate	28/03/2023	28/03/2023
									A071204	No Pacing	28/03/2023	28/03/2023
					CT233	Surgical	E0112	Dizziness	A071102	Failure to Interrogate	28/03/2023	28/03/2023
									A071204	No Pacing	28/03/2023	28/03/2023
							E0601	Arrhythmia	A071102	Failure to Interrogate	28/03/2023	28/03/2023
									A071204	No Pacing	28/03/2023	28/03/2023
							F1901	Additional Surgery	A071102	Failure to Interrogate	28/03/2023	28/03/2023
									A071204	No Pacing	28/03/2023	28/03/2023
					CT2406	Implantable	E0112	Dizziness	A071102	Failure to Interrogate	28/03/2023	28/03/2023
									A071204	No Pacing	28/03/2023	28/03/2023
							E0601	Arrhythmia	A071102	Failure to Interrogate	28/03/2023	28/03/2023
									A071204	No Pacing	28/03/2023	28/03/2023
							F1901	Additional Surgery	A071102	Failure to Interrogate	28/03/2023	28/03/2023
									A071204	No Pacing	28/03/2023	28/03/2023
					CT2407	Active implantable	E0112	Dizziness	A071102	Failure to Interrogate	28/03/2023	28/03/2023
									A071204	No Pacing	28/03/2023	28/03/2023
							E0601	Arrhythmia	A071102	Failure to Interrogate	28/03/2023	28/03/2023
									A071204	No Pacing	28/03/2023	28/03/2023
							F1901	Additional Surgery	A071102	Failure to Interrogate	28/03/2023	28/03/2023
									A071204	No Pacing	28/03/2023	28/03/2023
					CT257	Implantable cardiac pacemakers	E0112	Dizziness	A071102	Failure to Interrogate	28/03/2023	28/03/2023
									A071204	No Pacing	28/03/2023	28/03/2023
							E0601	Arrhythmia	A071102	Failure to Interrogate	28/03/2023	28/03/2023
									A071204	No Pacing	28/03/2023	28/03/2023

							F1901	Additional Surgery	A071102	Failure to Interrogate	28/03/2023	28/03/2023
									A071204	No Pacing	28/03/2023	28/03/2023
					CT2727	Active implant devices/ancillaries	E0112	Dizziness	A071102	Failure to Interrogate	28/03/2023	28/03/2023
									A071204	No Pacing	28/03/2023	28/03/2023
							E0601	Arrhythmia	A071102	Failure to Interrogate	28/03/2023	28/03/2023
									A071204	No Pacing	28/03/2023	28/03/2023
							F1901	Additional Surgery	A071102	Failure to Interrogate	28/03/2023	28/03/2023
									A071204	No Pacing	28/03/2023	28/03/2023
					CT2750	Implantable cardiac electrical stimulators and associated devices	E0112	Dizziness	A071102	Failure to Interrogate	28/03/2023	28/03/2023
									A071204	No Pacing	28/03/2023	28/03/2023
							E0601	Arrhythmia	A071102	Failure to Interrogate	28/03/2023	28/03/2023
									A071204	No Pacing	28/03/2023	28/03/2023
							F1901	Additional Surgery	A071102	Failure to Interrogate	28/03/2023	28/03/2023
									A071204	No Pacing	28/03/2023	28/03/2023
					CT2866	Implantable cardiac electrical stimulators	E0112	Dizziness	A071102	Failure to Interrogate	28/03/2023	28/03/2023
									A071204	No Pacing	28/03/2023	28/03/2023
							E0601	Arrhythmia	A071102	Failure to Interrogate	28/03/2023	28/03/2023
									A071204	No Pacing	28/03/2023	28/03/2023
							F1901	Additional Surgery	A071102	Failure to Interrogate	28/03/2023	28/03/2023
									A071204	No Pacing	28/03/2023	28/03/2023
					CT2985	Active	E0112	Dizziness	A071102	Failure to Interrogate	28/03/2023	28/03/2023
									A071204	No Pacing	28/03/2023	28/03/2023
							E0601	Arrhythmia	A071102	Failure to Interrogate	28/03/2023	28/03/2023
									A071204	No Pacing	28/03/2023	28/03/2023
							F1901	Additional Surgery	A071102	Failure to Interrogate	28/03/2023	28/03/2023
									A071204	No Pacing	28/03/2023	28/03/2023
					CT3001	Implantable metal	E0112	Dizziness	A071102	Failure to Interrogate	28/03/2023	28/03/2023
									A071204	No Pacing	28/03/2023	28/03/2023
							E0601	Arrhythmia	A071102	Failure to Interrogate	28/03/2023	28/03/2023
									A071204	No Pacing	28/03/2023	28/03/2023
							F1901	Additional Surgery	A071102	Failure to Interrogate	28/03/2023	28/03/2023
									A071204	No Pacing	28/03/2023	28/03/2023
					CT3002	Implantable synthetic polymer	E0112	Dizziness	A071102	Failure to Interrogate	28/03/2023	28/03/2023
									A071204	No Pacing	28/03/2023	28/03/2023
							E0601	Arrhythmia	A071102	Failure to Interrogate	28/03/2023	28/03/2023
									A071204	No Pacing	28/03/2023	28/03/2023
							F1901	Additional Surgery	A071102	Failure to Interrogate	28/03/2023	28/03/2023
									A071204	No Pacing	28/03/2023	28/03/2023
					CT3006	Implantable inorganic material	E0112	Dizziness	A071102	Failure to Interrogate	28/03/2023	28/03/2023
									A071204	No Pacing	28/03/2023	28/03/2023
							E0601	Arrhythmia	A071102	Failure to Interrogate	28/03/2023	28/03/2023
									A071204	No Pacing	28/03/2023	28/03/2023
							F1901	Additional Surgery	A071102	Failure to Interrogate	28/03/2023	28/03/2023
									A071204	No Pacing	28/03/2023	28/03/2023
					CT314	Electrically powered	E0112	Dizziness	A071102	Failure to Interrogate	28/03/2023	28/03/2023
									A071204	No Pacing	28/03/2023	28/03/2023
							E0601	Arrhythmia	A071102	Failure to Interrogate	28/03/2023	28/03/2023
									A071204	No Pacing	28/03/2023	28/03/2023
							F1901	Additional Surgery	A071102	Failure to Interrogate	28/03/2023	28/03/2023
									A071204	No Pacing	28/03/2023	28/03/2023
					CT321	Long-term surgical invasive	E0112	Dizziness	A071102	Failure to Interrogate	28/03/2023	28/03/2023
									A071204	No Pacing	28/03/2023	28/03/2023
							E0601	Arrhythmia	A071102	Failure to Interrogate	28/03/2023	28/03/2023
									A071204	No Pacing	28/03/2023	28/03/2023
							F1901	Additional Surgery	A071102	Failure to Interrogate	28/03/2023	28/03/2023
									A071204	No Pacing	28/03/2023	28/03/2023
					CT449	Generators and associated devices	E0112	Dizziness	A071102	Failure to Interrogate	28/03/2023	28/03/2023
									A071204	No Pacing	28/03/2023	28/03/2023
							E0601	Arrhythmia	A071102	Failure to Interrogate	28/03/2023	28/03/2023
									A071204	No Pacing	28/03/2023	28/03/2023
							F1901	Additional Surgery	A071102	Failure to Interrogate	28/03/2023	28/03/2023
									A071204	No Pacing	28/03/2023	28/03/2023
					CT981	Single use	E0112	Dizziness	A071102	Failure to Interrogate	28/03/2023	28/03/2023
									A071204	No Pacing	28/03/2023	28/03/2023
							E0601	Arrhythmia	A071102	Failure to Interrogate	28/03/2023	28/03/2023
									A071204	No Pacing	28/03/2023	28/03/2023
							F1901	Additional Surgery	A071102	Failure to Interrogate	28/03/2023	28/03/2023
									A071204	No Pacing	28/03/2023	28/03/2023
					CT983	Surgical invasive	E0112	Dizziness	A071102	Failure to Interrogate	28/03/2023	28/03/2023
									A071204	No Pacing	28/03/2023	28/03/2023
							E0601	Arrhythmia	A071102	Failure to Interrogate	28/03/2023	28/03/2023
									A071204	No Pacing	28/03/2023	28/03/2023
							F1901	Additional Surgery	A071102	Failure to Interrogate	28/03/2023	28/03/2023
									A071204	No Pacing	28/03/2023	28/03/2023
12716	2023	All other reportable incidents	46841	Cardiac pulse generator analyser cable, reusable	CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A0508	Noise, Audible		
									A070504	Premature Discharge of Battery		
									A0908	Incorrect, Inadequate or Imprecise Result or Readings		
							F24	Insufficient Health Impact Information	A0508	Noise, Audible		
									A070504	Premature Discharge of Battery		
									A0908	Incorrect, Inadequate or Imprecise Result or Readings		
12717	2023	All other reportable incidents	35223	Endocardial/interventricular septal pacing lead	CT1684	Cardiac pacemakers and associated devices	E0112	Dizziness	A070101	Failure to Capture		
									A072201	High Impedance		
							E2312	Fatigue	A070101	Failure to Capture		
									A072201	High Impedance		
							F1901	Additional Surgery	A070101	Failure to Capture		
									A072201	High Impedance		

12718	2023	All other reportable incidents	17882	Physiologic-monitoring defibrillation system	CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A0719	Unexpected Shutdown		
							F27	Problem Identified During Non-Clinical Procedure	A0719	Unexpected Shutdown		
12719	2023		47265	Dual-chamber implantable pacemaker, rate-responsive	CT141	Electrodes/Leads and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	08/01/2023	08/01/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	08/01/2023	08/01/2023
					CT145	Cardiovascular devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	08/01/2023	08/01/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	08/01/2023	08/01/2023
					CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	08/01/2023	08/01/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	08/01/2023	08/01/2023
					CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	08/01/2023	08/01/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	08/01/2023	08/01/2023
					CT233	Surgical	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	08/01/2023	08/01/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	08/01/2023	08/01/2023
					CT2406	Implantable	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	08/01/2023	08/01/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	08/01/2023	08/01/2023
					CT2407	Active implantable	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	08/01/2023	08/01/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	08/01/2023	08/01/2023
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	08/01/2023	08/01/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	08/01/2023	08/01/2023
					CT2727	Active implant devices/ancillaries	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	08/01/2023	08/01/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	08/01/2023	08/01/2023
					CT2750	Implantable cardiac electrical stimulators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	08/01/2023	08/01/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	08/01/2023	08/01/2023
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	08/01/2023	08/01/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	08/01/2023	08/01/2023
					CT2985	Active	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	08/01/2023	08/01/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	08/01/2023	08/01/2023
					CT3001	Implantable metal	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	08/01/2023	08/01/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	08/01/2023	08/01/2023
					CT3002	Implantable synthetic polymer	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	08/01/2023	08/01/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	08/01/2023	08/01/2023
					CT3006	Implantable inorganic material	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	08/01/2023	08/01/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	08/01/2023	08/01/2023
					CT314	Electrically powered	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	08/01/2023	08/01/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	08/01/2023	08/01/2023
					CT321	Long-term surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	08/01/2023	08/01/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	08/01/2023	08/01/2023
					CT449	Generators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	08/01/2023	08/01/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	08/01/2023	08/01/2023
					CT981	Single use	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	08/01/2023	08/01/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	08/01/2023	08/01/2023
					CT983	Surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	08/01/2023	08/01/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	08/01/2023	08/01/2023
12720	2023	All other reportable incidents	17882	Physiologic-monitoring defibrillation system	CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070908	Failure to Sense		
							F27	Problem Identified During Non-Clinical Procedure	A070908	Failure to Sense		
12721	2023		47265	Dual-chamber implantable pacemaker, rate-responsive	CT141	Electrodes/Leads and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	09/03/2023	09/03/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	09/03/2023	09/03/2023
					CT145	Cardiovascular devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	09/03/2023	09/03/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	09/03/2023	09/03/2023
					CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	09/03/2023	09/03/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	09/03/2023	09/03/2023
					CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	09/03/2023	09/03/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	09/03/2023	09/03/2023
					CT233	Surgical	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	09/03/2023	09/03/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	09/03/2023	09/03/2023
					CT2406	Implantable	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	09/03/2023	09/03/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	09/03/2023	09/03/2023
					CT2407	Active implantable	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	09/03/2023	09/03/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	09/03/2023	09/03/2023
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	09/03/2023	09/03/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	09/03/2023	09/03/2023
					CT2727	Active implant devices/ancillaries	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	09/03/2023	09/03/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	09/03/2023	09/03/2023
					CT2750	Implantable cardiac electrical stimulators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	09/03/2023	09/03/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	09/03/2023	09/03/2023
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	09/03/2023	09/03/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	09/03/2023	09/03/2023
					CT2985	Active	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	09/03/2023	09/03/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	09/03/2023	09/03/2023
					CT3001	Implantable metal	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	09/03/2023	09/03/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	09/03/2023	09/03/2023
					CT3002	Implantable synthetic polymer	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	09/03/2023	09/03/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	09/03/2023	09/03/2023
					CT3006	Implantable inorganic material	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	09/03/2023	09/03/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	09/03/2023	09/03/2023
					CT314	Electrically powered	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	09/03/2023	09/03/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	09/03/2023	09/03/2023
					CT321	Long-term surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	09/03/2023	09/03/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	09/03/2023	09/03/2023
					CT449	Generators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	09/03/2023	09/03/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	09/03/2023	09/03/2023
					CT981	Single use	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	09/03/2023	09/03/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	09/03/2023	09/03/2023
					CT983	Surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	09/03/2023	09/03/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	09/03/2023	09/03/2023
12722	2023	All other reportable incidents	47265	Dual-chamber implantable pacemaker, rate-responsive	CT141	Electrodes/Leads and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	26/02/2023	26/02/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	26/02/2023	26/02/2023
					CT145	Cardiovascular devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	26/02/2023	26/02/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	26/02/2023	26/02/2023

					CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	26/02/2023	26/02/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	26/02/2023	26/02/2023
					CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	26/02/2023	26/02/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	26/02/2023	26/02/2023
					CT233	Surgical	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	26/02/2023	26/02/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	26/02/2023	26/02/2023
					CT2406	Implantable	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	26/02/2023	26/02/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	26/02/2023	26/02/2023
					CT2407	Active implantable	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	26/02/2023	26/02/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	26/02/2023	26/02/2023
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	26/02/2023	26/02/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	26/02/2023	26/02/2023
					CT2727	Active implant devices/ancillaries	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	26/02/2023	26/02/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	26/02/2023	26/02/2023
					CT2750	Implantable cardiac electrical stimulators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	26/02/2023	26/02/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	26/02/2023	26/02/2023
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	26/02/2023	26/02/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	26/02/2023	26/02/2023
					CT2985	Active	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	26/02/2023	26/02/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	26/02/2023	26/02/2023
					CT3001	Implantable metal	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	26/02/2023	26/02/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	26/02/2023	26/02/2023
					CT3002	Implantable synthetic polymer	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	26/02/2023	26/02/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	26/02/2023	26/02/2023
					CT3006	Implantable inorganic material	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	26/02/2023	26/02/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	26/02/2023	26/02/2023
					CT314	Electrically powered	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	26/02/2023	26/02/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	26/02/2023	26/02/2023
					CT321	Long-term surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	26/02/2023	26/02/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	26/02/2023	26/02/2023
					CT449	Generators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	26/02/2023	26/02/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	26/02/2023	26/02/2023
					CT981	Single use	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	26/02/2023	26/02/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	26/02/2023	26/02/2023
					CT983	Surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	26/02/2023	26/02/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	26/02/2023	26/02/2023
12723	2023	All other reportable incidents	47265	Dual-chamber implantable pacemaker, rate-responsive	CT141	Electrodes/Leads and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	26/02/2023	26/02/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	26/02/2023	26/02/2023
					CT145	Cardiovascular devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	26/02/2023	26/02/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	26/02/2023	26/02/2023
					CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	26/02/2023	26/02/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	26/02/2023	26/02/2023
					CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	26/02/2023	26/02/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	26/02/2023	26/02/2023
					CT233	Surgical	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	26/02/2023	26/02/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	26/02/2023	26/02/2023
					CT2406	Implantable	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	26/02/2023	26/02/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	26/02/2023	26/02/2023
					CT2407	Active implantable	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	26/02/2023	26/02/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	26/02/2023	26/02/2023
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	26/02/2023	26/02/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	26/02/2023	26/02/2023
					CT2727	Active implant devices/ancillaries	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	26/02/2023	26/02/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	26/02/2023	26/02/2023
					CT2750	Implantable cardiac electrical stimulators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	26/02/2023	26/02/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	26/02/2023	26/02/2023
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	26/02/2023	26/02/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	26/02/2023	26/02/2023
					CT2985	Active	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	26/02/2023	26/02/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	26/02/2023	26/02/2023
					CT3001	Implantable metal	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	26/02/2023	26/02/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	26/02/2023	26/02/2023
					CT3002	Implantable synthetic polymer	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	26/02/2023	26/02/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	26/02/2023	26/02/2023
					CT3006	Implantable inorganic material	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	26/02/2023	26/02/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	26/02/2023	26/02/2023
					CT314	Electrically powered	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	26/02/2023	26/02/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	26/02/2023	26/02/2023
					CT321	Long-term surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	26/02/2023	26/02/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	26/02/2023	26/02/2023
					CT449	Generators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	26/02/2023	26/02/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	26/02/2023	26/02/2023
					CT981	Single use	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	26/02/2023	26/02/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	26/02/2023	26/02/2023
					CT983	Surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	26/02/2023	26/02/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	26/02/2023	26/02/2023
12724	2023	All other reportable incidents	47265	Dual-chamber implantable pacemaker, rate-responsive	CT141	Electrodes/Leads and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
					CT145	Cardiovascular devices	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
					CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		

					CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
					CT233	Surgical	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
					CT2406	Implantable	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
					CT2407	Active implantable	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
					CT2727	Active implant devices/ancillaries	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
					CT2750	Implantable cardiac electrical stimulators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
					CT2985	Active	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
					CT3001	Implantable metal	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
					CT3002	Implantable synthetic polymer	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
					CT3006	Implantable inorganic material	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
					CT314	Electrically powered	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
					CT321	Long-term surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
					CT449	Generators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
					CT981	Single use	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
					CT983	Surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
12725	2023	All other reportable incidents	17882	Physiologic-monitoring defibrillation system	CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070803	Failure to Power Up		
							F27	Problem Identified During Non-Clinical Procedure	A070803	Failure to Power Up		
12726	2023		47265	Dual-chamber implantable pacemaker, rate-responsive	CT141	Electrodes/Leads and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	26/02/2023	26/02/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	26/02/2023	26/02/2023
					CT145	Cardiovascular devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	26/02/2023	26/02/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	26/02/2023	26/02/2023
					CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	26/02/2023	26/02/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	26/02/2023	26/02/2023
					CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	26/02/2023	26/02/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	26/02/2023	26/02/2023
					CT233	Surgical	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	26/02/2023	26/02/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	26/02/2023	26/02/2023
					CT2406	Implantable	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	26/02/2023	26/02/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	26/02/2023	26/02/2023
					CT2407	Active implantable	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	26/02/2023	26/02/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	26/02/2023	26/02/2023
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	26/02/2023	26/02/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	26/02/2023	26/02/2023
					CT2727	Active implant devices/ancillaries	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	26/02/2023	26/02/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	26/02/2023	26/02/2023

					CT2750	Implantable cardiac electrical stimulators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	26/02/2023	26/02/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	26/02/2023	26/02/2023
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	26/02/2023	26/02/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	26/02/2023	26/02/2023
					CT2985	Active	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	26/02/2023	26/02/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	26/02/2023	26/02/2023
					CT3001	Implantable metal	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	26/02/2023	26/02/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	26/02/2023	26/02/2023
					CT3002	Implantable synthetic polymer	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	26/02/2023	26/02/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	26/02/2023	26/02/2023
					CT3006	Implantable inorganic material	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	26/02/2023	26/02/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	26/02/2023	26/02/2023
					CT314	Electrically powered	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	26/02/2023	26/02/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	26/02/2023	26/02/2023
					CT321	Long-term surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	26/02/2023	26/02/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	26/02/2023	26/02/2023
					CT449	Generators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	26/02/2023	26/02/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	26/02/2023	26/02/2023
					CT981	Single use	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	26/02/2023	26/02/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	26/02/2023	26/02/2023
					CT983	Surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	26/02/2023	26/02/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	26/02/2023	26/02/2023
12727	2023	All other reportable incidents	47265	Dual-chamber implantable pacemaker, rate-responsive	CT141	Electrodes/Leads and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	26/02/2023	26/02/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	26/02/2023	26/02/2023
					CT145	Cardiovascular devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	26/02/2023	26/02/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	26/02/2023	26/02/2023
					CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	26/02/2023	26/02/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	26/02/2023	26/02/2023
					CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	26/02/2023	26/02/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	26/02/2023	26/02/2023
					CT233	Surgical	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	26/02/2023	26/02/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	26/02/2023	26/02/2023
					CT2406	Implantable	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	26/02/2023	26/02/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	26/02/2023	26/02/2023
					CT2407	Active implantable	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	26/02/2023	26/02/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	26/02/2023	26/02/2023
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	26/02/2023	26/02/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	26/02/2023	26/02/2023
					CT2727	Active implant devices/ancillaries	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	26/02/2023	26/02/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	26/02/2023	26/02/2023
					CT2750	Implantable cardiac electrical stimulators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	26/02/2023	26/02/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	26/02/2023	26/02/2023
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	26/02/2023	26/02/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	26/02/2023	26/02/2023
					CT2985	Active	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	26/02/2023	26/02/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	26/02/2023	26/02/2023
					CT3001	Implantable metal	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	26/02/2023	26/02/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	26/02/2023	26/02/2023
					CT3002	Implantable synthetic polymer	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	26/02/2023	26/02/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	26/02/2023	26/02/2023
					CT3006	Implantable inorganic material	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	26/02/2023	26/02/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	26/02/2023	26/02/2023
					CT314	Electrically powered	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	26/02/2023	26/02/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	26/02/2023	26/02/2023
					CT321	Long-term surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	26/02/2023	26/02/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	26/02/2023	26/02/2023
					CT449	Generators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	26/02/2023	26/02/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	26/02/2023	26/02/2023
					CT981	Single use	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	26/02/2023	26/02/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	26/02/2023	26/02/2023
					CT983	Surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	26/02/2023	26/02/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	26/02/2023	26/02/2023
12728	2023	All other reportable incidents	47265	Dual-chamber implantable pacemaker, rate-responsive	CT141	Electrodes/Leads and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	08/03/2023	08/03/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	08/03/2023	08/03/2023
					CT145	Cardiovascular devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	08/03/2023	08/03/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	08/03/2023	08/03/2023
					CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	08/03/2023	08/03/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	08/03/2023	08/03/2023
					CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	08/03/2023	08/03/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	08/03/2023	08/03/2023
					CT233	Surgical	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	08/03/2023	08/03/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	08/03/2023	08/03/2023
					CT2406	Implantable	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	08/03/2023	08/03/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	08/03/2023	08/03/2023
					CT2407	Active implantable	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	08/03/2023	08/03/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	08/03/2023	08/03/2023
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	08/03/2023	08/03/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	08/03/2023	08/03/2023
					CT2727	Active implant devices/ancillaries	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	08/03/2023	08/03/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	08/03/2023	08/03/2023
					CT2750	Implantable cardiac electrical stimulators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	08/03/2023	08/03/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	08/03/2023	08/03/2023
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	08/03/2023	08/03/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	08/03/2023	08/03/2023
					CT2985	Active	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	08/03/2023	08/03/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	08/03/2023	08/03/2023
					CT3001	Implantable metal	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	08/03/2023	08/03/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	08/03/2023	08/03/2023

					CT3002	Implantable synthetic polymer	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	08/03/2023	08/03/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	08/03/2023	08/03/2023
					CT3006	Implantable inorganic material	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	08/03/2023	08/03/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	08/03/2023	08/03/2023
					CT314	Electrically powered	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	08/03/2023	08/03/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	08/03/2023	08/03/2023
					CT321	Long-term surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	08/03/2023	08/03/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	08/03/2023	08/03/2023
					CT449	Generators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	08/03/2023	08/03/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	08/03/2023	08/03/2023
					CT981	Single use	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	08/03/2023	08/03/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	08/03/2023	08/03/2023
					CT983	Surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	08/03/2023	08/03/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	08/03/2023	08/03/2023
12729	2023	All other reportable incidents	47265	Dual-chamber implantable pacemaker, rate-responsive	CT141	Electrodes/Leads and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	09/01/2023	09/01/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	09/01/2023	09/01/2023
					CT145	Cardiovascular devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	09/01/2023	09/01/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	09/01/2023	09/01/2023
					CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	09/01/2023	09/01/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	09/01/2023	09/01/2023
					CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	09/01/2023	09/01/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	09/01/2023	09/01/2023
					CT233	Surgical	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	09/01/2023	09/01/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	09/01/2023	09/01/2023
					CT2406	Implantable	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	09/01/2023	09/01/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	09/01/2023	09/01/2023
					CT2407	Active implantable	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	09/01/2023	09/01/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	09/01/2023	09/01/2023
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	09/01/2023	09/01/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	09/01/2023	09/01/2023
					CT2727	Active implant devices/ancillaries	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	09/01/2023	09/01/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	09/01/2023	09/01/2023
					CT2750	Implantable cardiac electrical stimulators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	09/01/2023	09/01/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	09/01/2023	09/01/2023
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	09/01/2023	09/01/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	09/01/2023	09/01/2023
					CT2985	Active	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	09/01/2023	09/01/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	09/01/2023	09/01/2023
					CT3001	Implantable metal	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	09/01/2023	09/01/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	09/01/2023	09/01/2023
					CT3002	Implantable synthetic polymer	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	09/01/2023	09/01/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	09/01/2023	09/01/2023
					CT3006	Implantable inorganic material	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	09/01/2023	09/01/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	09/01/2023	09/01/2023
					CT314	Electrically powered	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	09/01/2023	09/01/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	09/01/2023	09/01/2023
					CT321	Long-term surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	09/01/2023	09/01/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	09/01/2023	09/01/2023
					CT449	Generators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	09/01/2023	09/01/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	09/01/2023	09/01/2023
					CT981	Single use	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	09/01/2023	09/01/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	09/01/2023	09/01/2023
					CT983	Surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	09/01/2023	09/01/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	09/01/2023	09/01/2023
12730	2023		47265	Dual-chamber implantable pacemaker, rate-responsive	CT141	Electrodes/Leads and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	12/02/2023	12/02/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	12/02/2023	12/02/2023
					CT145	Cardiovascular devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	12/02/2023	12/02/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	12/02/2023	12/02/2023
					CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	12/02/2023	12/02/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	12/02/2023	12/02/2023
					CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	12/02/2023	12/02/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	12/02/2023	12/02/2023
					CT233	Surgical	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	12/02/2023	12/02/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	12/02/2023	12/02/2023
					CT2406	Implantable	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	12/02/2023	12/02/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	12/02/2023	12/02/2023
					CT2407	Active implantable	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	12/02/2023	12/02/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	12/02/2023	12/02/2023
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	12/02/2023	12/02/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	12/02/2023	12/02/2023
					CT2727	Active implant devices/ancillaries	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	12/02/2023	12/02/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	12/02/2023	12/02/2023
					CT2750	Implantable cardiac electrical stimulators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	12/02/2023	12/02/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	12/02/2023	12/02/2023
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	12/02/2023	12/02/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	12/02/2023	12/02/2023
					CT2985	Active	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	12/02/2023	12/02/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	12/02/2023	12/02/2023
					CT3001	Implantable metal	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	12/02/2023	12/02/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	12/02/2023	12/02/2023
					CT3002	Implantable synthetic polymer	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	12/02/2023	12/02/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	12/02/2023	12/02/2023
					CT3006	Implantable inorganic material	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	12/02/2023	12/02/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	12/02/2023	12/02/2023
					CT314	Electrically powered	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	12/02/2023	12/02/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	12/02/2023	12/02/2023
					CT321	Long-term surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	12/02/2023	12/02/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	12/02/2023	12/02/2023

					CT449	Generators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	12/02/2023	12/02/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	12/02/2023	12/02/2023
					CT981	Single use	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	12/02/2023	12/02/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	12/02/2023	12/02/2023
					CT983	Surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	12/02/2023	12/02/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	12/02/2023	12/02/2023
12731	2023	All other reportable incidents	17882	Physiologic-monitoring defibrillation system	CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A0719	Unexpected Shutdown		
							F26	No Health Consequences or Impact	A0719	Unexpected Shutdown		
12732	2023	All other reportable incidents	47270	Cardiac resynchronization therapy implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F12	Serious Injury/ Illness/ Impairment	A070504	Premature Discharge of Battery		
							F23	Unexpected Medical Intervention	A070504	Premature Discharge of Battery		
12733	2023		47265	Dual-chamber implantable pacemaker, rate-responsive	CT141	Electrodes/Leads and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	07/03/2023	07/03/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	07/03/2023	07/03/2023
					CT145	Cardiovascular devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	07/03/2023	07/03/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	07/03/2023	07/03/2023
					CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	07/03/2023	07/03/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	07/03/2023	07/03/2023
					CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	07/03/2023	07/03/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	07/03/2023	07/03/2023
					CT233	Surgical	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	07/03/2023	07/03/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	07/03/2023	07/03/2023
					CT2406	Implantable	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	07/03/2023	07/03/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	07/03/2023	07/03/2023
					CT2407	Active implantable	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	07/03/2023	07/03/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	07/03/2023	07/03/2023
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	07/03/2023	07/03/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	07/03/2023	07/03/2023
					CT2727	Active implant devices/ancillaries	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	07/03/2023	07/03/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	07/03/2023	07/03/2023
					CT2750	Implantable cardiac electrical stimulators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	07/03/2023	07/03/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	07/03/2023	07/03/2023
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	07/03/2023	07/03/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	07/03/2023	07/03/2023
					CT2985	Active	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	07/03/2023	07/03/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	07/03/2023	07/03/2023
					CT3001	Implantable metal	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	07/03/2023	07/03/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	07/03/2023	07/03/2023
					CT3002	Implantable synthetic polymer	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	07/03/2023	07/03/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	07/03/2023	07/03/2023
					CT3006	Implantable inorganic material	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	07/03/2023	07/03/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	07/03/2023	07/03/2023
					CT314	Electrically powered	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	07/03/2023	07/03/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	07/03/2023	07/03/2023
					CT321	Long-term surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	07/03/2023	07/03/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	07/03/2023	07/03/2023
					CT449	Generators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	07/03/2023	07/03/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	07/03/2023	07/03/2023
					CT981	Single use	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	07/03/2023	07/03/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	07/03/2023	07/03/2023
					CT983	Surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	07/03/2023	07/03/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	07/03/2023	07/03/2023
12734	2023	All other reportable incidents	47265	Dual-chamber implantable pacemaker, rate-responsive	CT141	Electrodes/Leads and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	17/03/2023	17/03/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	17/03/2023	17/03/2023
					CT145	Cardiovascular devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	17/03/2023	17/03/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	17/03/2023	17/03/2023
					CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	17/03/2023	17/03/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	17/03/2023	17/03/2023
					CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	17/03/2023	17/03/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	17/03/2023	17/03/2023
					CT233	Surgical	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	17/03/2023	17/03/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	17/03/2023	17/03/2023
					CT2406	Implantable	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	17/03/2023	17/03/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	17/03/2023	17/03/2023
					CT2407	Active implantable	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	17/03/2023	17/03/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	17/03/2023	17/03/2023
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	17/03/2023	17/03/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	17/03/2023	17/03/2023
					CT2727	Active implant devices/ancillaries	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	17/03/2023	17/03/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	17/03/2023	17/03/2023
					CT2750	Implantable cardiac electrical stimulators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	17/03/2023	17/03/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	17/03/2023	17/03/2023
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	17/03/2023	17/03/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	17/03/2023	17/03/2023
					CT2985	Active	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	17/03/2023	17/03/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	17/03/2023	17/03/2023
					CT3001	Implantable metal	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	17/03/2023	17/03/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	17/03/2023	17/03/2023
					CT3002	Implantable synthetic polymer	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	17/03/2023	17/03/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	17/03/2023	17/03/2023
					CT3006	Implantable inorganic material	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	17/03/2023	17/03/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	17/03/2023	17/03/2023
					CT314	Electrically powered	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	17/03/2023	17/03/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	17/03/2023	17/03/2023
					CT321	Long-term surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	17/03/2023	17/03/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	17/03/2023	17/03/2023
					CT449	Generators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	17/03/2023	17/03/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	17/03/2023	17/03/2023
					CT981	Single use	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	17/03/2023	17/03/2023

							F1901	Additional Surgery	A24	No Apparent Device Problem	17/03/2023	17/03/2023
					CT983	Surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	17/03/2023	17/03/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	17/03/2023	17/03/2023
12735	2023	All other reportable incidents	17882	Physiologic-monitoring defibrillation system	CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070908	Failure to Sense		
							F26	No Health Consequences or Impact	A070908	Failure to Sense		
12736	2023	All other reportable incidents	60910	Coronary venous pacing lead	CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A072201	High Impedance		
									A1502	Positioning Problem		
							F26	No Health Consequences or Impact	A072201	High Impedance		
									A1502	Positioning Problem		
12737	2023	All other reportable incidents	37265	Dual-chamber implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A072201	High Impedance		
									A0908	Incorrect, Inadequate or Imprecise Result or Readings		
							F26	No Health Consequences or Impact	A072201	High Impedance		
									A0908	Incorrect, Inadequate or Imprecise Result or Readings		
12738	2023	All other reportable incidents	36241	Pacemaker/defibrillator lead	CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A04	Material Integrity Problem		
									A072201	High Impedance		
							F26	No Health Consequences or Impact	A04	Material Integrity Problem		
									A072201	High Impedance		
12739	2023	All other reportable incidents	47265	Dual-chamber implantable pacemaker, rate-responsive	CT141	Electrodes/Leads and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	24/02/2023	24/02/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	24/02/2023	24/02/2023
					CT145	Cardiovascular devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	24/02/2023	24/02/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	24/02/2023	24/02/2023
					CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	24/02/2023	24/02/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	24/02/2023	24/02/2023
					CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	24/02/2023	24/02/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	24/02/2023	24/02/2023
					CT233	Surgical	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	24/02/2023	24/02/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	24/02/2023	24/02/2023
					CT2406	Implantable	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	24/02/2023	24/02/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	24/02/2023	24/02/2023
					CT2407	Active implantable	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	24/02/2023	24/02/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	24/02/2023	24/02/2023
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	24/02/2023	24/02/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	24/02/2023	24/02/2023
					CT2727	Active implant devices/ancillaries	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	24/02/2023	24/02/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	24/02/2023	24/02/2023
					CT2750	Implantable cardiac electrical stimulators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	24/02/2023	24/02/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	24/02/2023	24/02/2023
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	24/02/2023	24/02/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	24/02/2023	24/02/2023
					CT2985	Active	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	24/02/2023	24/02/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	24/02/2023	24/02/2023
					CT3001	Implantable metal	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	24/02/2023	24/02/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	24/02/2023	24/02/2023
					CT3002	Implantable synthetic polymer	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	24/02/2023	24/02/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	24/02/2023	24/02/2023
					CT3006	Implantable inorganic material	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	24/02/2023	24/02/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	24/02/2023	24/02/2023
					CT314	Electrically powered	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	24/02/2023	24/02/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	24/02/2023	24/02/2023
					CT321	Long-term surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	24/02/2023	24/02/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	24/02/2023	24/02/2023
					CT449	Generators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	24/02/2023	24/02/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	24/02/2023	24/02/2023
					CT981	Single use	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	24/02/2023	24/02/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	24/02/2023	24/02/2023
					CT983	Surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	24/02/2023	24/02/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	24/02/2023	24/02/2023
12740	2023	All other reportable incidents	47265	Dual-chamber implantable pacemaker, rate-responsive	CT141	Electrodes/Leads and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	02/03/2023	02/03/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	02/03/2023	02/03/2023
					CT145	Cardiovascular devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	02/03/2023	02/03/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	02/03/2023	02/03/2023
					CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	02/03/2023	02/03/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	02/03/2023	02/03/2023
					CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	02/03/2023	02/03/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	02/03/2023	02/03/2023
					CT233	Surgical	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	02/03/2023	02/03/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	02/03/2023	02/03/2023
					CT2406	Implantable	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	02/03/2023	02/03/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	02/03/2023	02/03/2023
					CT2407	Active implantable	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	02/03/2023	02/03/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	02/03/2023	02/03/2023
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	02/03/2023	02/03/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	02/03/2023	02/03/2023
					CT2727	Active implant devices/ancillaries	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	02/03/2023	02/03/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	02/03/2023	02/03/2023
					CT2750	Implantable cardiac electrical stimulators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	02/03/2023	02/03/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	02/03/2023	02/03/2023
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	02/03/2023	02/03/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	02/03/2023	02/03/2023
					CT2985	Active	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	02/03/2023	02/03/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	02/03/2023	02/03/2023
					CT3001	Implantable metal	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	02/03/2023	02/03/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	02/03/2023	02/03/2023
					CT3002	Implantable synthetic polymer	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	02/03/2023	02/03/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	02/03/2023	02/03/2023
					CT3006	Implantable inorganic material	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	02/03/2023	02/03/2023

							F1901	Additional Surgery	A24	No Apparent Device Problem	02/03/2023	02/03/2023
					CT314	Electrically powered	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	02/03/2023	02/03/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	02/03/2023	02/03/2023
					CT321	Long-term surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	02/03/2023	02/03/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	02/03/2023	02/03/2023
					CT449	Generators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	02/03/2023	02/03/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	02/03/2023	02/03/2023
					CT981	Single use	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	02/03/2023	02/03/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	02/03/2023	02/03/2023
					CT983	Surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	02/03/2023	02/03/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	02/03/2023	02/03/2023
12741	2023	All other reportable incidents	47265	Dual-chamber implantable pacemaker, rate-responsive	CT141	Electrodes/Leads and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A0908	Incorrect, Inadequate or Imprecise Result or Readings	20/03/2023	20/03/2023
							F08	Hospitalization or Prolonged Hospitalization	A0908	Incorrect, Inadequate or Imprecise Result or Readings	20/03/2023	20/03/2023
							F1905	Device Revision or Replacement	A0908	Incorrect, Inadequate or Imprecise Result or Readings	20/03/2023	20/03/2023
					CT145	Cardiovascular devices	E2403	No Clinical Signs, Symptoms or Conditions	A0908	Incorrect, Inadequate or Imprecise Result or Readings	20/03/2023	20/03/2023
							F08	Hospitalization or Prolonged Hospitalization	A0908	Incorrect, Inadequate or Imprecise Result or Readings	20/03/2023	20/03/2023
							F1905	Device Revision or Replacement	A0908	Incorrect, Inadequate or Imprecise Result or Readings	20/03/2023	20/03/2023
					CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A0908	Incorrect, Inadequate or Imprecise Result or Readings	20/03/2023	20/03/2023
							F08	Hospitalization or Prolonged Hospitalization	A0908	Incorrect, Inadequate or Imprecise Result or Readings	20/03/2023	20/03/2023
							F1905	Device Revision or Replacement	A0908	Incorrect, Inadequate or Imprecise Result or Readings	20/03/2023	20/03/2023
					CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A0908	Incorrect, Inadequate or Imprecise Result or Readings	20/03/2023	20/03/2023
							F08	Hospitalization or Prolonged Hospitalization	A0908	Incorrect, Inadequate or Imprecise Result or Readings	20/03/2023	20/03/2023
							F1905	Device Revision or Replacement	A0908	Incorrect, Inadequate or Imprecise Result or Readings	20/03/2023	20/03/2023
					CT233	Surgical	E2403	No Clinical Signs, Symptoms or Conditions	A0908	Incorrect, Inadequate or Imprecise Result or Readings	20/03/2023	20/03/2023
							F08	Hospitalization or Prolonged Hospitalization	A0908	Incorrect, Inadequate or Imprecise Result or Readings	20/03/2023	20/03/2023
							F1905	Device Revision or Replacement	A0908	Incorrect, Inadequate or Imprecise Result or Readings	20/03/2023	20/03/2023
					CT2406	Implantable	E2403	No Clinical Signs, Symptoms or Conditions	A0908	Incorrect, Inadequate or Imprecise Result or Readings	20/03/2023	20/03/2023
							F08	Hospitalization or Prolonged Hospitalization	A0908	Incorrect, Inadequate or Imprecise Result or Readings	20/03/2023	20/03/2023
							F1905	Device Revision or Replacement	A0908	Incorrect, Inadequate or Imprecise Result or Readings	20/03/2023	20/03/2023
					CT2407	Active implantable	E2403	No Clinical Signs, Symptoms or Conditions	A0908	Incorrect, Inadequate or Imprecise Result or Readings	20/03/2023	20/03/2023
							F08	Hospitalization or Prolonged Hospitalization	A0908	Incorrect, Inadequate or Imprecise Result or Readings	20/03/2023	20/03/2023
							F1905	Device Revision or Replacement	A0908	Incorrect, Inadequate or Imprecise Result or Readings	20/03/2023	20/03/2023
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A0908	Incorrect, Inadequate or Imprecise Result or Readings	20/03/2023	20/03/2023
							F08	Hospitalization or Prolonged Hospitalization	A0908	Incorrect, Inadequate or Imprecise Result or Readings	20/03/2023	20/03/2023
							F1905	Device Revision or Replacement	A0908	Incorrect, Inadequate or Imprecise Result or Readings	20/03/2023	20/03/2023
					CT2727	Active implant devices/ancillaries	E2403	No Clinical Signs, Symptoms or Conditions	A0908	Incorrect, Inadequate or Imprecise Result or Readings	20/03/2023	20/03/2023
							F08	Hospitalization or Prolonged Hospitalization	A0908	Incorrect, Inadequate or Imprecise Result or Readings	20/03/2023	20/03/2023
							F1905	Device Revision or Replacement	A0908	Incorrect, Inadequate or Imprecise Result or Readings	20/03/2023	20/03/2023
					CT2750	Implantable cardiac electrical stimulators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A0908	Incorrect, Inadequate or Imprecise Result or Readings	20/03/2023	20/03/2023
							F08	Hospitalization or Prolonged Hospitalization	A0908	Incorrect, Inadequate or Imprecise Result or Readings	20/03/2023	20/03/2023
							F1905	Device Revision or Replacement	A0908	Incorrect, Inadequate or Imprecise Result or Readings	20/03/2023	20/03/2023
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A0908	Incorrect, Inadequate or Imprecise Result or Readings	20/03/2023	20/03/2023
							F08	Hospitalization or Prolonged Hospitalization	A0908	Incorrect, Inadequate or Imprecise Result or Readings	20/03/2023	20/03/2023
							F1905	Device Revision or Replacement	A0908	Incorrect, Inadequate or Imprecise Result or Readings	20/03/2023	20/03/2023
					CT2985	Active	E2403	No Clinical Signs, Symptoms or Conditions	A0908	Incorrect, Inadequate or Imprecise Result or Readings	20/03/2023	20/03/2023
							F08	Hospitalization or Prolonged Hospitalization	A0908	Incorrect, Inadequate or Imprecise Result or Readings	20/03/2023	20/03/2023
							F1905	Device Revision or Replacement	A0908	Incorrect, Inadequate or Imprecise Result or Readings	20/03/2023	20/03/2023
					CT3001	Implantable metal	E2403	No Clinical Signs, Symptoms or Conditions	A0908	Incorrect, Inadequate or Imprecise Result or Readings	20/03/2023	20/03/2023
							F08	Hospitalization or Prolonged Hospitalization	A0908	Incorrect, Inadequate or Imprecise Result or Readings	20/03/2023	20/03/2023
							F1905	Device Revision or Replacement	A0908	Incorrect, Inadequate or Imprecise Result or Readings	20/03/2023	20/03/2023
					CT3002	Implantable synthetic polymer	E2403	No Clinical Signs, Symptoms or Conditions	A0908	Incorrect, Inadequate or Imprecise Result or Readings	20/03/2023	20/03/2023
							F08	Hospitalization or Prolonged Hospitalization	A0908	Incorrect, Inadequate or Imprecise Result or Readings	20/03/2023	20/03/2023
							F1905	Device Revision or Replacement	A0908	Incorrect, Inadequate or Imprecise Result or Readings	20/03/2023	20/03/2023
					CT3006	Implantable inorganic material	E2403	No Clinical Signs, Symptoms or Conditions	A0908	Incorrect, Inadequate or Imprecise Result or Readings	20/03/2023	20/03/2023

							F08	Hospitalization or Prolonged Hospitalization	A0908	Incorrect, Inadequate or Imprecise Result or Readings	20/03/2023	20/03/2023
							F1905	Device Revision or Replacement	A0908	Incorrect, Inadequate or Imprecise Result or Readings	20/03/2023	20/03/2023
					CT314	Electrically powered	E2403	No Clinical Signs, Symptoms or Conditions	A0908	Incorrect, Inadequate or Imprecise Result or Readings	20/03/2023	20/03/2023
							F08	Hospitalization or Prolonged Hospitalization	A0908	Incorrect, Inadequate or Imprecise Result or Readings	20/03/2023	20/03/2023
							F1905	Device Revision or Replacement	A0908	Incorrect, Inadequate or Imprecise Result or Readings	20/03/2023	20/03/2023
					CT321	Long-term surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A0908	Incorrect, Inadequate or Imprecise Result or Readings	20/03/2023	20/03/2023
							F08	Hospitalization or Prolonged Hospitalization	A0908	Incorrect, Inadequate or Imprecise Result or Readings	20/03/2023	20/03/2023
							F1905	Device Revision or Replacement	A0908	Incorrect, Inadequate or Imprecise Result or Readings	20/03/2023	20/03/2023
					CT449	Generators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A0908	Incorrect, Inadequate or Imprecise Result or Readings	20/03/2023	20/03/2023
							F08	Hospitalization or Prolonged Hospitalization	A0908	Incorrect, Inadequate or Imprecise Result or Readings	20/03/2023	20/03/2023
							F1905	Device Revision or Replacement	A0908	Incorrect, Inadequate or Imprecise Result or Readings	20/03/2023	20/03/2023
					CT981	Single use	E2403	No Clinical Signs, Symptoms or Conditions	A0908	Incorrect, Inadequate or Imprecise Result or Readings	20/03/2023	20/03/2023
							F08	Hospitalization or Prolonged Hospitalization	A0908	Incorrect, Inadequate or Imprecise Result or Readings	20/03/2023	20/03/2023
							F1905	Device Revision or Replacement	A0908	Incorrect, Inadequate or Imprecise Result or Readings	20/03/2023	20/03/2023
					CT983	Surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A0908	Incorrect, Inadequate or Imprecise Result or Readings	20/03/2023	20/03/2023
							F08	Hospitalization or Prolonged Hospitalization	A0908	Incorrect, Inadequate or Imprecise Result or Readings	20/03/2023	20/03/2023
							F1905	Device Revision or Replacement	A0908	Incorrect, Inadequate or Imprecise Result or Readings	20/03/2023	20/03/2023
12742	2023	All other reportable incidents	17882	Physiologic-monitoring defibrillation system	CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A1603	Failure of Device to Self-Test		
							F27	Problem Identified During Non-Clinical Procedure	A1603	Failure of Device to Self-Test		
12743	2023	All other reportable incidents	47263	Cardiac resynchronization therapy implantable pacemaker	CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A0908	Incorrect, Inadequate or Imprecise Result or Readings	22/03/2023	22/03/2023
									A1304	Telemetry Discrepancy	22/03/2023	22/03/2023
							F08	Hospitalization or Prolonged Hospitalization	A0908	Incorrect, Inadequate or Imprecise Result or Readings	22/03/2023	22/03/2023
									A1304	Telemetry Discrepancy	22/03/2023	22/03/2023
							F1905	Device Revision or Replacement	A0908	Incorrect, Inadequate or Imprecise Result or Readings	22/03/2023	22/03/2023
									A1304	Telemetry Discrepancy	22/03/2023	22/03/2023
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A0908	Incorrect, Inadequate or Imprecise Result or Readings	22/03/2023	22/03/2023
									A1304	Telemetry Discrepancy	22/03/2023	22/03/2023
							F08	Hospitalization or Prolonged Hospitalization	A0908	Incorrect, Inadequate or Imprecise Result or Readings	22/03/2023	22/03/2023
									A1304	Telemetry Discrepancy	22/03/2023	22/03/2023
							F1905	Device Revision or Replacement	A0908	Incorrect, Inadequate or Imprecise Result or Readings	22/03/2023	22/03/2023
									A1304	Telemetry Discrepancy	22/03/2023	22/03/2023
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A0908	Incorrect, Inadequate or Imprecise Result or Readings	22/03/2023	22/03/2023
									A1304	Telemetry Discrepancy	22/03/2023	22/03/2023
							F08	Hospitalization or Prolonged Hospitalization	A0908	Incorrect, Inadequate or Imprecise Result or Readings	22/03/2023	22/03/2023
									A1304	Telemetry Discrepancy	22/03/2023	22/03/2023
							F1905	Device Revision or Replacement	A0908	Incorrect, Inadequate or Imprecise Result or Readings	22/03/2023	22/03/2023
									A1304	Telemetry Discrepancy	22/03/2023	22/03/2023
12744	2023	All other reportable incidents	35852	Single-chamber implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A1304	Telemetry Discrepancy		
							F26	No Health Consequences or Impact	A1304	Telemetry Discrepancy		
12745	2023	All other reportable incidents	47263	Cardiac resynchronization therapy implantable pacemaker	CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A0908	Incorrect, Inadequate or Imprecise Result or Readings	09/03/2023	09/03/2023
							F08	Hospitalization or Prolonged Hospitalization	A0908	Incorrect, Inadequate or Imprecise Result or Readings	09/03/2023	09/03/2023
							F1905	Device Revision or Replacement	A0908	Incorrect, Inadequate or Imprecise Result or Readings	09/03/2023	09/03/2023
					CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A0908	Incorrect, Inadequate or Imprecise Result or Readings	09/03/2023	09/03/2023
							F08	Hospitalization or Prolonged Hospitalization	A0908	Incorrect, Inadequate or Imprecise Result or Readings	09/03/2023	09/03/2023
							F1905	Device Revision or Replacement	A0908	Incorrect, Inadequate or Imprecise Result or Readings	09/03/2023	09/03/2023
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A0908	Incorrect, Inadequate or Imprecise Result or Readings	09/03/2023	09/03/2023
							F08	Hospitalization or Prolonged Hospitalization	A0908	Incorrect, Inadequate or Imprecise Result or Readings	09/03/2023	09/03/2023
							F1905	Device Revision or Replacement	A0908	Incorrect, Inadequate or Imprecise Result or Readings	09/03/2023	09/03/2023
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A0908	Incorrect, Inadequate or Imprecise Result or Readings	09/03/2023	09/03/2023
							F08	Hospitalization or Prolonged Hospitalization	A0908	Incorrect, Inadequate or Imprecise Result or Readings	09/03/2023	09/03/2023
							F1905	Device Revision or Replacement	A0908	Incorrect, Inadequate or Imprecise Result or Readings	09/03/2023	09/03/2023
12746	2023		47265	Dual-chamber implantable pacemaker, rate-responsive	CT141	Electrodes/Leads and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	10/03/2023	10/03/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	10/03/2023	10/03/2023
					CT145	Cardiovascular devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	10/03/2023	10/03/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	10/03/2023	10/03/2023
					CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	10/03/2023	10/03/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	10/03/2023	10/03/2023
					CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	10/03/2023	10/03/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	10/03/2023	10/03/2023

					CT233	Surgical	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	10/03/2023	10/03/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	10/03/2023	10/03/2023
					CT2406	Implantable	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	10/03/2023	10/03/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	10/03/2023	10/03/2023
					CT2407	Active implantable	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	10/03/2023	10/03/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	10/03/2023	10/03/2023
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	10/03/2023	10/03/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	10/03/2023	10/03/2023
					CT2727	Active implant devices/ancillaries	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	10/03/2023	10/03/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	10/03/2023	10/03/2023
					CT2750	Implantable cardiac electrical stimulators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	10/03/2023	10/03/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	10/03/2023	10/03/2023
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	10/03/2023	10/03/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	10/03/2023	10/03/2023
					CT2985	Active	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	10/03/2023	10/03/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	10/03/2023	10/03/2023
					CT3001	Implantable metal	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	10/03/2023	10/03/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	10/03/2023	10/03/2023
					CT3002	Implantable synthetic polymer	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	10/03/2023	10/03/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	10/03/2023	10/03/2023
					CT3006	Implantable inorganic material	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	10/03/2023	10/03/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	10/03/2023	10/03/2023
					CT314	Electrically powered	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	10/03/2023	10/03/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	10/03/2023	10/03/2023
					CT321	Long-term surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	10/03/2023	10/03/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	10/03/2023	10/03/2023
					CT449	Generators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	10/03/2023	10/03/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	10/03/2023	10/03/2023
					CT981	Single use	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	10/03/2023	10/03/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	10/03/2023	10/03/2023
					CT983	Surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	10/03/2023	10/03/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	10/03/2023	10/03/2023
12747	2023		47265	Dual-chamber implantable pacemaker, rate-responsive	CT141	Electrodes/Leads and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	02/03/2023	02/03/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	02/03/2023	02/03/2023
					CT145	Cardiovascular devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	02/03/2023	02/03/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	02/03/2023	02/03/2023
					CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	02/03/2023	02/03/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	02/03/2023	02/03/2023
					CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	02/03/2023	02/03/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	02/03/2023	02/03/2023
					CT233	Surgical	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	02/03/2023	02/03/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	02/03/2023	02/03/2023
					CT2406	Implantable	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	02/03/2023	02/03/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	02/03/2023	02/03/2023
					CT2407	Active implantable	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	02/03/2023	02/03/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	02/03/2023	02/03/2023
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	02/03/2023	02/03/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	02/03/2023	02/03/2023
					CT2727	Active implant devices/ancillaries	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	02/03/2023	02/03/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	02/03/2023	02/03/2023
					CT2750	Implantable cardiac electrical stimulators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	02/03/2023	02/03/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	02/03/2023	02/03/2023
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	02/03/2023	02/03/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	02/03/2023	02/03/2023
					CT2985	Active	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	02/03/2023	02/03/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	02/03/2023	02/03/2023
					CT3001	Implantable metal	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	02/03/2023	02/03/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	02/03/2023	02/03/2023
					CT3002	Implantable synthetic polymer	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	02/03/2023	02/03/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	02/03/2023	02/03/2023
					CT3006	Implantable inorganic material	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	02/03/2023	02/03/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	02/03/2023	02/03/2023
					CT314	Electrically powered	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	02/03/2023	02/03/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	02/03/2023	02/03/2023
					CT321	Long-term surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	02/03/2023	02/03/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	02/03/2023	02/03/2023
					CT449	Generators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	02/03/2023	02/03/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	02/03/2023	02/03/2023
					CT981	Single use	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	02/03/2023	02/03/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	02/03/2023	02/03/2023
					CT983	Surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	02/03/2023	02/03/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	02/03/2023	02/03/2023
12748	2023		47265	Dual-chamber implantable pacemaker, rate-responsive	CT141	Electrodes/Leads and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	20/03/2023	20/03/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	20/03/2023	20/03/2023
					CT145	Cardiovascular devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	20/03/2023	20/03/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	20/03/2023	20/03/2023
					CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	20/03/2023	20/03/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	20/03/2023	20/03/2023
					CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	20/03/2023	20/03/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	20/03/2023	20/03/2023
					CT233	Surgical	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	20/03/2023	20/03/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	20/03/2023	20/03/2023
					CT2406	Implantable	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	20/03/2023	20/03/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	20/03/2023	20/03/2023
					CT2407	Active implantable	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	20/03/2023	20/03/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	20/03/2023	20/03/2023
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	20/03/2023	20/03/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	20/03/2023	20/03/2023

					CT2727	Active implant devices/ancillaries	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	20/03/2023	20/03/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	20/03/2023	20/03/2023
					CT2750	Implantable cardiac electrical stimulators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	20/03/2023	20/03/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	20/03/2023	20/03/2023
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	20/03/2023	20/03/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	20/03/2023	20/03/2023
					CT2985	Active	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	20/03/2023	20/03/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	20/03/2023	20/03/2023
					CT3001	Implantable metal	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	20/03/2023	20/03/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	20/03/2023	20/03/2023
					CT3002	Implantable synthetic polymer	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	20/03/2023	20/03/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	20/03/2023	20/03/2023
					CT3006	Implantable inorganic material	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	20/03/2023	20/03/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	20/03/2023	20/03/2023
					CT314	Electrically powered	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	20/03/2023	20/03/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	20/03/2023	20/03/2023
					CT321	Long-term surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	20/03/2023	20/03/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	20/03/2023	20/03/2023
					CT449	Generators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	20/03/2023	20/03/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	20/03/2023	20/03/2023
					CT981	Single use	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	20/03/2023	20/03/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	20/03/2023	20/03/2023
					CT983	Surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	20/03/2023	20/03/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	20/03/2023	20/03/2023
12749	2023		47265	Dual-chamber implantable pacemaker, rate-responsive	CT141	Electrodes/Leads and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	03/02/2023	03/02/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	03/02/2023	03/02/2023
					CT145	Cardiovascular devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	03/02/2023	03/02/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	03/02/2023	03/02/2023
					CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	03/02/2023	03/02/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	03/02/2023	03/02/2023
					CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	03/02/2023	03/02/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	03/02/2023	03/02/2023
					CT233	Surgical	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	03/02/2023	03/02/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	03/02/2023	03/02/2023
					CT2406	Implantable	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	03/02/2023	03/02/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	03/02/2023	03/02/2023
					CT2407	Active implantable	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	03/02/2023	03/02/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	03/02/2023	03/02/2023
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	03/02/2023	03/02/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	03/02/2023	03/02/2023
					CT2727	Active implant devices/ancillaries	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	03/02/2023	03/02/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	03/02/2023	03/02/2023
					CT2750	Implantable cardiac electrical stimulators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	03/02/2023	03/02/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	03/02/2023	03/02/2023
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	03/02/2023	03/02/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	03/02/2023	03/02/2023
					CT2985	Active	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	03/02/2023	03/02/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	03/02/2023	03/02/2023
					CT3001	Implantable metal	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	03/02/2023	03/02/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	03/02/2023	03/02/2023
					CT3002	Implantable synthetic polymer	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	03/02/2023	03/02/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	03/02/2023	03/02/2023
					CT3006	Implantable inorganic material	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	03/02/2023	03/02/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	03/02/2023	03/02/2023
					CT314	Electrically powered	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	03/02/2023	03/02/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	03/02/2023	03/02/2023
					CT321	Long-term surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	03/02/2023	03/02/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	03/02/2023	03/02/2023
					CT449	Generators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	03/02/2023	03/02/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	03/02/2023	03/02/2023
					CT981	Single use	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	03/02/2023	03/02/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	03/02/2023	03/02/2023
					CT983	Surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	03/02/2023	03/02/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	03/02/2023	03/02/2023
12750	2023	All other reportable incidents	35223	Endocardial/interventricular septal pacing lead	CT1684	Cardiac pacemakers and associated devices	E0601	Arrhythmia	A051201	Device Dislodged or Dislocated		
									A070101	Failure to Capture		
							F1901	Additional Surgery	A051201	Device Dislodged or Dislocated		
									A070101	Failure to Capture		
12751	2023	All other reportable incidents	35223	Endocardial/interventricular septal pacing lead	CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A040101	Fracture		
							F1908	Prolonged Surgery	A040101	Fracture		
12752	2023	All other reportable incidents	35852	Single-chamber implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
12753	2023		47265	Dual-chamber implantable pacemaker, rate-responsive	CT141	Electrodes/Leads and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	10/03/2023	10/03/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	10/03/2023	10/03/2023
					CT145	Cardiovascular devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	10/03/2023	10/03/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	10/03/2023	10/03/2023
					CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	10/03/2023	10/03/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	10/03/2023	10/03/2023
					CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	10/03/2023	10/03/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	10/03/2023	10/03/2023
					CT233	Surgical	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	10/03/2023	10/03/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	10/03/2023	10/03/2023
					CT2406	Implantable	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	10/03/2023	10/03/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	10/03/2023	10/03/2023
					CT2407	Active implantable	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	10/03/2023	10/03/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	10/03/2023	10/03/2023
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	10/03/2023	10/03/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	10/03/2023	10/03/2023

					CT2727	Active implant devices/ancillaries	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	10/03/2023	10/03/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	10/03/2023	10/03/2023
					CT2750	Implantable cardiac electrical stimulators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	10/03/2023	10/03/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	10/03/2023	10/03/2023
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	10/03/2023	10/03/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	10/03/2023	10/03/2023
					CT2985	Active	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	10/03/2023	10/03/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	10/03/2023	10/03/2023
					CT3001	Implantable metal	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	10/03/2023	10/03/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	10/03/2023	10/03/2023
					CT3002	Implantable synthetic polymer	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	10/03/2023	10/03/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	10/03/2023	10/03/2023
					CT3006	Implantable inorganic material	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	10/03/2023	10/03/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	10/03/2023	10/03/2023
					CT314	Electrically powered	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	10/03/2023	10/03/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	10/03/2023	10/03/2023
					CT321	Long-term surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	10/03/2023	10/03/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	10/03/2023	10/03/2023
					CT449	Generators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	10/03/2023	10/03/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	10/03/2023	10/03/2023
					CT981	Single use	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	10/03/2023	10/03/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	10/03/2023	10/03/2023
					CT983	Surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	10/03/2023	10/03/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	10/03/2023	10/03/2023
12754	2023		47265	Dual-chamber implantable pacemaker, rate-responsive	CT141	Electrodes/Leads and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	03/03/2023	03/03/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	03/03/2023	03/03/2023
					CT145	Cardiovascular devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	03/03/2023	03/03/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	03/03/2023	03/03/2023
					CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	03/03/2023	03/03/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	03/03/2023	03/03/2023
					CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	03/03/2023	03/03/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	03/03/2023	03/03/2023
					CT233	Surgical	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	03/03/2023	03/03/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	03/03/2023	03/03/2023
					CT2406	Implantable	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	03/03/2023	03/03/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	03/03/2023	03/03/2023
					CT2407	Active implantable	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	03/03/2023	03/03/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	03/03/2023	03/03/2023
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	03/03/2023	03/03/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	03/03/2023	03/03/2023
					CT2727	Active implant devices/ancillaries	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	03/03/2023	03/03/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	03/03/2023	03/03/2023
					CT2750	Implantable cardiac electrical stimulators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	03/03/2023	03/03/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	03/03/2023	03/03/2023
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	03/03/2023	03/03/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	03/03/2023	03/03/2023
					CT2985	Active	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	03/03/2023	03/03/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	03/03/2023	03/03/2023
					CT3001	Implantable metal	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	03/03/2023	03/03/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	03/03/2023	03/03/2023
					CT3002	Implantable synthetic polymer	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	03/03/2023	03/03/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	03/03/2023	03/03/2023
					CT3006	Implantable inorganic material	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	03/03/2023	03/03/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	03/03/2023	03/03/2023
					CT314	Electrically powered	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	03/03/2023	03/03/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	03/03/2023	03/03/2023
					CT321	Long-term surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	03/03/2023	03/03/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	03/03/2023	03/03/2023
					CT449	Generators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	03/03/2023	03/03/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	03/03/2023	03/03/2023
					CT981	Single use	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	03/03/2023	03/03/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	03/03/2023	03/03/2023
					CT983	Surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	03/03/2023	03/03/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	03/03/2023	03/03/2023
12755	2023	All other reportable incidents	35852	Single-chamber implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	22/09/2022	22/09/2022
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	22/09/2022	22/09/2022
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	22/09/2022	22/09/2022
12756	2023	All other reportable incidents	35852	Single-chamber implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem	26/10/2022	26/10/2022
									A070504	Premature Discharge of Battery	26/10/2022	26/10/2022
							F08	Hospitalization or Prolonged Hospitalization	A0705	Battery Problem	26/10/2022	26/10/2022
									A070504	Premature Discharge of Battery	26/10/2022	26/10/2022
							F1903	Device Explantation	A0705	Battery Problem	26/10/2022	26/10/2022
									A070504	Premature Discharge of Battery	26/10/2022	26/10/2022
12757	2023	All other reportable incidents	35852	Single-chamber implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
12758	2023	All other reportable incidents	35852	Single-chamber implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E2104	Electric Shock	A0709	Device Sensing Problem		
									A070909	Oversensing		
									A071302	Inappropriate/Inadequate Shock/Stimulation		
							F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A0709	Device Sensing Problem		
									A070909	Oversensing		
									A071302	Inappropriate/Inadequate Shock/Stimulation		
							F2203	Imaging Required	A0709	Device Sensing Problem		
									A070909	Oversensing		
									A071302	Inappropriate/Inadequate Shock/Stimulation		
12759	2023	All other reportable incidents	47265	Dual-chamber implantable pacemaker, rate-responsive	CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	13/03/2023	13/03/2023
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	13/03/2023	13/03/2023
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	13/03/2023	13/03/2023
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	13/03/2023	13/03/2023

							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	13/03/2023	13/03/2023
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	13/03/2023	13/03/2023
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	13/03/2023	13/03/2023
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	13/03/2023	13/03/2023
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	13/03/2023	13/03/2023
12760	2023	All other reportable incidents	35223	Endocardial/interventricular septal pacing lead	CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070909	Oversensing		
							F26	No Health Consequences or Impact	A070909	Oversensing		
12761	2023	All other reportable incidents	47265	Dual-chamber implantable pacemaker, rate-responsive	CT141	Electrodes/Leads and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	28/04/2022	28/04/2022
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery	28/04/2022	28/04/2022
					CT145	Cardiovascular devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	28/04/2022	28/04/2022
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery	28/04/2022	28/04/2022
					CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	28/04/2022	28/04/2022
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery	28/04/2022	28/04/2022
					CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	28/04/2022	28/04/2022
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery	28/04/2022	28/04/2022
					CT233	Surgical	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	28/04/2022	28/04/2022
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery	28/04/2022	28/04/2022
					CT2406	Implantable	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	28/04/2022	28/04/2022
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery	28/04/2022	28/04/2022
					CT2407	Active implantable	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	28/04/2022	28/04/2022
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery	28/04/2022	28/04/2022
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	28/04/2022	28/04/2022
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery	28/04/2022	28/04/2022
					CT2727	Active implant devices/ancillaries	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	28/04/2022	28/04/2022
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery	28/04/2022	28/04/2022
					CT2750	Implantable cardiac electrical stimulators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	28/04/2022	28/04/2022
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery	28/04/2022	28/04/2022
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	28/04/2022	28/04/2022
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery	28/04/2022	28/04/2022
					CT2985	Active	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	28/04/2022	28/04/2022
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery	28/04/2022	28/04/2022
					CT3001	Implantable metal	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	28/04/2022	28/04/2022
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery	28/04/2022	28/04/2022
					CT3002	Implantable synthetic polymer	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	28/04/2022	28/04/2022
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery	28/04/2022	28/04/2022
					CT3006	Implantable inorganic material	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	28/04/2022	28/04/2022
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery	28/04/2022	28/04/2022
					CT314	Electrically powered	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	28/04/2022	28/04/2022
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery	28/04/2022	28/04/2022
					CT321	Long-term surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	28/04/2022	28/04/2022
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery	28/04/2022	28/04/2022
					CT449	Generators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	28/04/2022	28/04/2022
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery	28/04/2022	28/04/2022
					CT981	Single use	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	28/04/2022	28/04/2022
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery	28/04/2022	28/04/2022
					CT983	Surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	28/04/2022	28/04/2022
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery	28/04/2022	28/04/2022
12762	2023	All other reportable incidents	47265	Dual-chamber implantable pacemaker, rate-responsive	CT141	Electrodes/Leads and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A23	Use of Device Problem		
							F26	No Health Consequences or Impact	A23	Use of Device Problem		
					CT145	Cardiovascular devices	E2403	No Clinical Signs, Symptoms or Conditions	A23	Use of Device Problem		
							F26	No Health Consequences or Impact	A23	Use of Device Problem		
					CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A23	Use of Device Problem		
							F26	No Health Consequences or Impact	A23	Use of Device Problem		
					CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A23	Use of Device Problem		
							F26	No Health Consequences or Impact	A23	Use of Device Problem		
					CT233	Surgical	E2403	No Clinical Signs, Symptoms or Conditions	A23	Use of Device Problem		
							F26	No Health Consequences or Impact	A23	Use of Device Problem		
					CT2406	Implantable	E2403	No Clinical Signs, Symptoms or Conditions	A23	Use of Device Problem		
							F26	No Health Consequences or Impact	A23	Use of Device Problem		
					CT2407	Active implantable	E2403	No Clinical Signs, Symptoms or Conditions	A23	Use of Device Problem		
							F26	No Health Consequences or Impact	A23	Use of Device Problem		
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A23	Use of Device Problem		
							F26	No Health Consequences or Impact	A23	Use of Device Problem		
					CT2727	Active implant devices/ancillaries	E2403	No Clinical Signs, Symptoms or Conditions	A23	Use of Device Problem		
							F26	No Health Consequences or Impact	A23	Use of Device Problem		
					CT2750	Implantable cardiac electrical stimulators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A23	Use of Device Problem		
							F26	No Health Consequences or Impact	A23	Use of Device Problem		
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A23	Use of Device Problem		
							F26	No Health Consequences or Impact	A23	Use of Device Problem		
					CT2985	Active	E2403	No Clinical Signs, Symptoms or Conditions	A23	Use of Device Problem		
							F26	No Health Consequences or Impact	A23	Use of Device Problem		
					CT3001	Implantable metal	E2403	No Clinical Signs, Symptoms or Conditions	A23	Use of Device Problem		
							F26	No Health Consequences or Impact	A23	Use of Device Problem		
					CT3002	Implantable synthetic polymer	E2403	No Clinical Signs, Symptoms or Conditions	A23	Use of Device Problem		
							F26	No Health Consequences or Impact	A23	Use of Device Problem		
					CT3006	Implantable inorganic material	E2403	No Clinical Signs, Symptoms or Conditions	A23	Use of Device Problem		
							F26	No Health Consequences or Impact	A23	Use of Device Problem		
					CT314	Electrically powered	E2403	No Clinical Signs, Symptoms or Conditions	A23	Use of Device Problem		
							F26	No Health Consequences or Impact	A23	Use of Device Problem		
					CT321	Long-term surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A23	Use of Device Problem		
							F26	No Health Consequences or Impact	A23	Use of Device Problem		
					CT449	Generators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A23	Use of Device Problem		
							F26	No Health Consequences or Impact	A23	Use of Device Problem		
					CT981	Single use	E2403	No Clinical Signs, Symptoms or Conditions	A23	Use of Device Problem		
							F26	No Health Consequences or Impact	A23	Use of Device Problem		
					CT983	Surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A23	Use of Device Problem		
							F26	No Health Consequences or Impact	A23	Use of Device Problem		
12763	2023				CT257	Implantable cardiac pacemakers	F02	Death	A0701	Capturing Problem		

12764	2023		47265	Dual-chamber implantable pacemaker, rate-responsive	CT141	Electrodes/Leads and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem		
							F1901	Additional Surgery	A24	No Apparent Device Problem		
					CT145	Cardiovascular devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem		
							F1901	Additional Surgery	A24	No Apparent Device Problem		
					CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem		
							F1901	Additional Surgery	A24	No Apparent Device Problem		
					CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem		
							F1901	Additional Surgery	A24	No Apparent Device Problem		
					CT233	Surgical	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem		
							F1901	Additional Surgery	A24	No Apparent Device Problem		
					CT2406	Implantable	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem		
							F1901	Additional Surgery	A24	No Apparent Device Problem		
					CT2407	Active implantable	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem		
							F1901	Additional Surgery	A24	No Apparent Device Problem		
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem		
							F1901	Additional Surgery	A24	No Apparent Device Problem		
					CT2727	Active implant devices/ancillaries	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem		
							F1901	Additional Surgery	A24	No Apparent Device Problem		
					CT2750	Implantable cardiac electrical stimulators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem		
							F1901	Additional Surgery	A24	No Apparent Device Problem		
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem		
							F1901	Additional Surgery	A24	No Apparent Device Problem		
					CT2985	Active	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem		
							F1901	Additional Surgery	A24	No Apparent Device Problem		
					CT3001	Implantable metal	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem		
							F1901	Additional Surgery	A24	No Apparent Device Problem		
					CT3002	Implantable synthetic polymer	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem		
							F1901	Additional Surgery	A24	No Apparent Device Problem		
					CT3006	Implantable inorganic material	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem		
							F1901	Additional Surgery	A24	No Apparent Device Problem		
					CT314	Electrically powered	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem		
							F1901	Additional Surgery	A24	No Apparent Device Problem		
					CT321	Long-term surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem		
							F1901	Additional Surgery	A24	No Apparent Device Problem		
					CT449	Generators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem		
							F1901	Additional Surgery	A24	No Apparent Device Problem		
					CT981	Single use	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem		
							F1901	Additional Surgery	A24	No Apparent Device Problem		
					CT983	Surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem		
							F1901	Additional Surgery	A24	No Apparent Device Problem		
12765	2023	All other reportable incidents	47265	Dual-chamber implantable pacemaker, rate-responsive	CT141	Electrodes/Leads and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT145	Cardiovascular devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT233	Surgical	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT2406	Implantable	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT2407	Active implantable	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT2727	Active implant devices/ancillaries	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT2750	Implantable cardiac electrical stimulators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT2985	Active	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT3001	Implantable metal	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT3002	Implantable synthetic polymer	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT3006	Implantable inorganic material	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT314	Electrically powered	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT321	Long-term surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT449	Generators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT981	Single use	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT983	Surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
12766	2023	All other reportable incidents	47267	Single-chamber implantable pacemaker, rate-responsive	CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	01/07/2023	01/07/2023
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	01/07/2023	01/07/2023
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	01/07/2023	01/07/2023
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	01/07/2023	01/07/2023
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	01/07/2023	01/07/2023
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	01/07/2023	01/07/2023
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	01/07/2023	01/07/2023
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	01/07/2023	01/07/2023

							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	01/07/2023	01/07/2023
12767	2023	All other reportable incidents	17882	Physiologic-monitoring defibrillation system	CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A0801	Failure to Calibrate		
							F27	Problem Identified During Non-Clinical Procedure	A0801	Failure to Calibrate		
12768	2023		35852	Single-chamber implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E17	Skin and Subcutaneous Tissue	A0705	Battery Problem		
							E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem		
							F26	No Health Consequences or Impact	A0705	Battery Problem		
12769	2023	All other reportable incidents	35223	Endocardial/interventricular septal pacing lead	CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070909	Oversensing		
							F26	No Health Consequences or Impact	A070909	Oversensing		
12770	2023	All other reportable incidents	60789	Intracardiac pacemaker	CT1685	Cardiac pacemakers	E2402	Appropriate Clinical Signs, Symptoms and Conditions Term/Code not Available	A070504	Premature Discharge of Battery		
							F12	Serious Injury/ Illness/ Impairment	A070504	Premature Discharge of Battery		
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery		
							F23	Unexpected Medical Intervention	A070504	Premature Discharge of Battery		
					CT257	Implantable cardiac pacemakers	E2402	Appropriate Clinical Signs, Symptoms and Conditions Term/Code not Available	A070504	Premature Discharge of Battery		
							F12	Serious Injury/ Illness/ Impairment	A070504	Premature Discharge of Battery		
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery		
							F23	Unexpected Medical Intervention	A070504	Premature Discharge of Battery		
					CT2818	Catheterization/Cannulation/Fluid path devices	E2402	Appropriate Clinical Signs, Symptoms and Conditions Term/Code not Available	A070504	Premature Discharge of Battery		
							F12	Serious Injury/ Illness/ Impairment	A070504	Premature Discharge of Battery		
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery		
							F23	Unexpected Medical Intervention	A070504	Premature Discharge of Battery		
					CT2866	Implantable cardiac electrical stimulators	E2402	Appropriate Clinical Signs, Symptoms and Conditions Term/Code not Available	A070504	Premature Discharge of Battery		
							F12	Serious Injury/ Illness/ Impairment	A070504	Premature Discharge of Battery		
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery		
							F23	Unexpected Medical Intervention	A070504	Premature Discharge of Battery		
12771	2023	All other reportable incidents	47267	Single-chamber implantable pacemaker, rate-responsive	CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A072201	High Impedance		
							F26	No Health Consequences or Impact	A072201	High Impedance		
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A072201	High Impedance		
							F26	No Health Consequences or Impact	A072201	High Impedance		
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A072201	High Impedance		
							F26	No Health Consequences or Impact	A072201	High Impedance		
12772	2023	All other reportable incidents	47267	Single-chamber implantable pacemaker, rate-responsive	CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
12773	2023	All other reportable incidents	47267	Single-chamber implantable pacemaker, rate-responsive	CT141	Electrodes/Leads and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT145	Cardiovascular devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT233	Surgical	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT2406	Implantable	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT2407	Active implantable	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT2727	Active implant devices/ancillaries	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT2750	Implantable cardiac electrical stimulators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT2985	Active	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT3001	Implantable metal	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT3002	Implantable synthetic polymer	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT3006	Implantable inorganic material	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT314	Electrically powered	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT321	Long-term surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT449	Generators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT981	Single use	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT983	Surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
12774	2023		47265	Dual-chamber implantable pacemaker, rate-responsive	CT141	Electrodes/Leads and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	27/01/2023	27/01/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	27/01/2023	27/01/2023
					CT145	Cardiovascular devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	27/01/2023	27/01/2023

							F1901	Additional Surgery	A24	No Apparent Device Problem	27/01/2023	27/01/2023
					CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	27/01/2023	27/01/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	27/01/2023	27/01/2023
					CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	27/01/2023	27/01/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	27/01/2023	27/01/2023
					CT233	Surgical	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	27/01/2023	27/01/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	27/01/2023	27/01/2023
					CT2406	Implantable	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	27/01/2023	27/01/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	27/01/2023	27/01/2023
					CT2407	Active implantable	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	27/01/2023	27/01/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	27/01/2023	27/01/2023
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	27/01/2023	27/01/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	27/01/2023	27/01/2023
					CT2727	Active implant devices/ancillaries	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	27/01/2023	27/01/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	27/01/2023	27/01/2023
					CT2750	Implantable cardiac electrical stimulators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	27/01/2023	27/01/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	27/01/2023	27/01/2023
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	27/01/2023	27/01/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	27/01/2023	27/01/2023
					CT2985	Active	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	27/01/2023	27/01/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	27/01/2023	27/01/2023
					CT3001	Implantable metal	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	27/01/2023	27/01/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	27/01/2023	27/01/2023
					CT3002	Implantable synthetic polymer	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	27/01/2023	27/01/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	27/01/2023	27/01/2023
					CT3006	Implantable inorganic material	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	27/01/2023	27/01/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	27/01/2023	27/01/2023
					CT314	Electrically powered	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	27/01/2023	27/01/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	27/01/2023	27/01/2023
					CT321	Long-term surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	27/01/2023	27/01/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	27/01/2023	27/01/2023
					CT449	Generators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	27/01/2023	27/01/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	27/01/2023	27/01/2023
					CT981	Single use	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	27/01/2023	27/01/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	27/01/2023	27/01/2023
					CT983	Surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	27/01/2023	27/01/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	27/01/2023	27/01/2023
12775	2023	All other reportable incidents	35852	Single-chamber implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
12776	2023		47265	Dual-chamber implantable pacemaker, rate-responsive	CT141	Electrodes/Leads and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	10/03/2023	10/03/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	10/03/2023	10/03/2023
					CT145	Cardiovascular devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	10/03/2023	10/03/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	10/03/2023	10/03/2023
					CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	10/03/2023	10/03/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	10/03/2023	10/03/2023
					CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	10/03/2023	10/03/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	10/03/2023	10/03/2023
					CT233	Surgical	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	10/03/2023	10/03/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	10/03/2023	10/03/2023
					CT2406	Implantable	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	10/03/2023	10/03/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	10/03/2023	10/03/2023
					CT2407	Active implantable	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	10/03/2023	10/03/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	10/03/2023	10/03/2023
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	10/03/2023	10/03/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	10/03/2023	10/03/2023
					CT2727	Active implant devices/ancillaries	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	10/03/2023	10/03/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	10/03/2023	10/03/2023
					CT2750	Implantable cardiac electrical stimulators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	10/03/2023	10/03/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	10/03/2023	10/03/2023
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	10/03/2023	10/03/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	10/03/2023	10/03/2023
					CT2985	Active	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	10/03/2023	10/03/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	10/03/2023	10/03/2023
					CT3001	Implantable metal	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	10/03/2023	10/03/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	10/03/2023	10/03/2023
					CT3002	Implantable synthetic polymer	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	10/03/2023	10/03/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	10/03/2023	10/03/2023
					CT3006	Implantable inorganic material	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	10/03/2023	10/03/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	10/03/2023	10/03/2023
					CT314	Electrically powered	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	10/03/2023	10/03/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	10/03/2023	10/03/2023
					CT321	Long-term surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	10/03/2023	10/03/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	10/03/2023	10/03/2023
					CT449	Generators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	10/03/2023	10/03/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	10/03/2023	10/03/2023
					CT981	Single use	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	10/03/2023	10/03/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	10/03/2023	10/03/2023
					CT983	Surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	10/03/2023	10/03/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	10/03/2023	10/03/2023
12777	2023		47265	Dual-chamber implantable pacemaker, rate-responsive	CT141	Electrodes/Leads and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	28/03/2023	28/03/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	28/03/2023	28/03/2023
					CT145	Cardiovascular devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	28/03/2023	28/03/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	28/03/2023	28/03/2023
					CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	28/03/2023	28/03/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	28/03/2023	28/03/2023
					CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	28/03/2023	28/03/2023

							F1901	Additional Surgery	A24	No Apparent Device Problem	28/03/2023	28/03/2023
					CT233	Surgical	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	28/03/2023	28/03/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	28/03/2023	28/03/2023
					CT2406	Implantable	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	28/03/2023	28/03/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	28/03/2023	28/03/2023
					CT2407	Active implantable	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	28/03/2023	28/03/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	28/03/2023	28/03/2023
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	28/03/2023	28/03/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	28/03/2023	28/03/2023
					CT2727	Active implant devices/ancillaries	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	28/03/2023	28/03/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	28/03/2023	28/03/2023
					CT2750	Implantable cardiac electrical stimulators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	28/03/2023	28/03/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	28/03/2023	28/03/2023
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	28/03/2023	28/03/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	28/03/2023	28/03/2023
					CT2985	Active	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	28/03/2023	28/03/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	28/03/2023	28/03/2023
					CT3001	Implantable metal	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	28/03/2023	28/03/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	28/03/2023	28/03/2023
					CT3002	Implantable synthetic polymer	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	28/03/2023	28/03/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	28/03/2023	28/03/2023
					CT3006	Implantable inorganic material	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	28/03/2023	28/03/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	28/03/2023	28/03/2023
					CT314	Electrically powered	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	28/03/2023	28/03/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	28/03/2023	28/03/2023
					CT321	Long-term surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	28/03/2023	28/03/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	28/03/2023	28/03/2023
					CT449	Generators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	28/03/2023	28/03/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	28/03/2023	28/03/2023
					CT981	Single use	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	28/03/2023	28/03/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	28/03/2023	28/03/2023
					CT983	Surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	28/03/2023	28/03/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	28/03/2023	28/03/2023
12778	2023		47265	Dual-chamber implantable pacemaker, rate-responsive	CT141	Electrodes/Leads and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	04/01/2023	04/01/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	04/01/2023	04/01/2023
					CT145	Cardiovascular devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	04/01/2023	04/01/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	04/01/2023	04/01/2023
					CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	04/01/2023	04/01/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	04/01/2023	04/01/2023
					CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	04/01/2023	04/01/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	04/01/2023	04/01/2023
					CT233	Surgical	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	04/01/2023	04/01/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	04/01/2023	04/01/2023
					CT2406	Implantable	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	04/01/2023	04/01/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	04/01/2023	04/01/2023
					CT2407	Active implantable	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	04/01/2023	04/01/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	04/01/2023	04/01/2023
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	04/01/2023	04/01/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	04/01/2023	04/01/2023
					CT2727	Active implant devices/ancillaries	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	04/01/2023	04/01/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	04/01/2023	04/01/2023
					CT2750	Implantable cardiac electrical stimulators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	04/01/2023	04/01/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	04/01/2023	04/01/2023
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	04/01/2023	04/01/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	04/01/2023	04/01/2023
					CT2985	Active	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	04/01/2023	04/01/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	04/01/2023	04/01/2023
					CT3001	Implantable metal	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	04/01/2023	04/01/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	04/01/2023	04/01/2023
					CT3002	Implantable synthetic polymer	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	04/01/2023	04/01/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	04/01/2023	04/01/2023
					CT3006	Implantable inorganic material	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	04/01/2023	04/01/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	04/01/2023	04/01/2023
					CT314	Electrically powered	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	04/01/2023	04/01/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	04/01/2023	04/01/2023
					CT321	Long-term surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	04/01/2023	04/01/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	04/01/2023	04/01/2023
					CT449	Generators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	04/01/2023	04/01/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	04/01/2023	04/01/2023
					CT981	Single use	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	04/01/2023	04/01/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	04/01/2023	04/01/2023
					CT983	Surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	04/01/2023	04/01/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	04/01/2023	04/01/2023
12779	2023		47265	Dual-chamber implantable pacemaker, rate-responsive	CT141	Electrodes/Leads and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	27/01/2023	27/01/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	27/01/2023	27/01/2023
					CT145	Cardiovascular devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	27/01/2023	27/01/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	27/01/2023	27/01/2023
					CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	27/01/2023	27/01/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	27/01/2023	27/01/2023
					CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	27/01/2023	27/01/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	27/01/2023	27/01/2023
					CT233	Surgical	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	27/01/2023	27/01/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	27/01/2023	27/01/2023
					CT2406	Implantable	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	27/01/2023	27/01/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	27/01/2023	27/01/2023
					CT2407	Active implantable	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	27/01/2023	27/01/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	27/01/2023	27/01/2023
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	27/01/2023	27/01/2023

							F1901	Additional Surgery	A24	No Apparent Device Problem	27/01/2023	27/01/2023
					CT2727	Active implant devices/ancillaries	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	27/01/2023	27/01/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	27/01/2023	27/01/2023
					CT2750	Implantable cardiac electrical stimulators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	27/01/2023	27/01/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	27/01/2023	27/01/2023
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	27/01/2023	27/01/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	27/01/2023	27/01/2023
					CT2985	Active	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	27/01/2023	27/01/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	27/01/2023	27/01/2023
					CT3001	Implantable metal	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	27/01/2023	27/01/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	27/01/2023	27/01/2023
					CT3002	Implantable synthetic polymer	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	27/01/2023	27/01/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	27/01/2023	27/01/2023
					CT3006	Implantable inorganic material	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	27/01/2023	27/01/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	27/01/2023	27/01/2023
					CT314	Electrically powered	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	27/01/2023	27/01/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	27/01/2023	27/01/2023
					CT321	Long-term surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	27/01/2023	27/01/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	27/01/2023	27/01/2023
					CT449	Generators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	27/01/2023	27/01/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	27/01/2023	27/01/2023
					CT981	Single use	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	27/01/2023	27/01/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	27/01/2023	27/01/2023
					CT983	Surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	27/01/2023	27/01/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	27/01/2023	27/01/2023
12780	2023	All other reportable incidents	17882	Physiologic-monitoring defibrillation system	CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070803	Failure to Power Up		
							F27	Problem Identified During Non-Clinical Procedure	A070803	Failure to Power Up		
12781	2023	All other reportable incidents	37265	Dual-chamber implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	29/03/2023	29/03/2023
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	29/03/2023	29/03/2023
							F12	Serious Injury/ Illness/ Impairment	A070504	Premature Discharge of Battery	29/03/2023	29/03/2023
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	29/03/2023	29/03/2023
12782	2023	All other reportable incidents	37265	Dual-chamber implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	14/04/2023	14/04/2023
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	14/04/2023	14/04/2023
							F12	Serious Injury/ Illness/ Impairment	A070504	Premature Discharge of Battery	14/04/2023	14/04/2023
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	14/04/2023	14/04/2023
12783	2023	All other reportable incidents	36241	Pacemaker/defibrillator lead	CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A072201	High Impedance		
							F26	No Health Consequences or Impact	A072201	High Impedance		
12784	2023	All other reportable incidents	35852	Single-chamber implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E2104	Electric Shock	A070909	Oversensing		
									A071302	Inappropriate/Inadequate Shock/Stimulation		
									A090801	Signal Artifact/Noise		
									A1301	Failure to Read Input Signal		
							F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A070909	Oversensing		
									A071302	Inappropriate/Inadequate Shock/Stimulation		
									A090801	Signal Artifact/Noise		
									A1301	Failure to Read Input Signal		
12785	2023	All other reportable incidents	35223	Endocardial/interventricular septal pacing lead	CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A0401	Break	24/03/2023	24/03/2023
									A070909	Oversensing	24/03/2023	24/03/2023
									A090801	Signal Artifact/Noise	24/03/2023	24/03/2023
							F1901	Additional Surgery	A0401	Break	24/03/2023	24/03/2023
									A070909	Oversensing	24/03/2023	24/03/2023
									A090801	Signal Artifact/Noise	24/03/2023	24/03/2023
12786	2023	All other reportable incidents	47265	Dual-chamber implantable pacemaker, rate-responsive	CT141	Electrodes/Leads and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	09/02/2023	09/02/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	09/02/2023	09/02/2023
					CT145	Cardiovascular devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	09/02/2023	09/02/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	09/02/2023	09/02/2023
					CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	09/02/2023	09/02/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	09/02/2023	09/02/2023
					CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	09/02/2023	09/02/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	09/02/2023	09/02/2023
					CT233	Surgical	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	09/02/2023	09/02/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	09/02/2023	09/02/2023
					CT2406	Implantable	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	09/02/2023	09/02/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	09/02/2023	09/02/2023
					CT2407	Active implantable	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	09/02/2023	09/02/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	09/02/2023	09/02/2023
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	09/02/2023	09/02/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	09/02/2023	09/02/2023
					CT2727	Active implant devices/ancillaries	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	09/02/2023	09/02/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	09/02/2023	09/02/2023
					CT2750	Implantable cardiac electrical stimulators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	09/02/2023	09/02/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	09/02/2023	09/02/2023
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	09/02/2023	09/02/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	09/02/2023	09/02/2023
					CT2985	Active	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	09/02/2023	09/02/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	09/02/2023	09/02/2023
					CT3001	Implantable metal	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	09/02/2023	09/02/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	09/02/2023	09/02/2023
					CT3002	Implantable synthetic polymer	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	09/02/2023	09/02/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	09/02/2023	09/02/2023
					CT3006	Implantable inorganic material	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	09/02/2023	09/02/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	09/02/2023	09/02/2023
					CT314	Electrically powered	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	09/02/2023	09/02/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	09/02/2023	09/02/2023
					CT321	Long-term surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	09/02/2023	09/02/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	09/02/2023	09/02/2023
					CT449	Generators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	09/02/2023	09/02/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	09/02/2023	09/02/2023
					CT981	Single use	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	09/02/2023	09/02/2023

						F1901	Additional Surgery	A24	No Apparent Device Problem	09/02/2023	09/02/2023
					CT983	Surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	09/02/2023	09/02/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	09/02/2023	09/02/2023
12787	2023		47265	Dual-chamber implantable pacemaker, rate-responsive	CT141	Electrodes/Leads and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	09/12/2022	09/12/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	09/12/2022	09/12/2022
					CT145	Cardiovascular devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	09/12/2022	09/12/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	09/12/2022	09/12/2022
					CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	09/12/2022	09/12/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	09/12/2022	09/12/2022
					CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	09/12/2022	09/12/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	09/12/2022	09/12/2022
					CT233	Surgical	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	09/12/2022	09/12/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	09/12/2022	09/12/2022
					CT2406	Implantable	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	09/12/2022	09/12/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	09/12/2022	09/12/2022
					CT2407	Active implantable	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	09/12/2022	09/12/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	09/12/2022	09/12/2022
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	09/12/2022	09/12/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	09/12/2022	09/12/2022
					CT2727	Active implant devices/ancillaries	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	09/12/2022	09/12/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	09/12/2022	09/12/2022
					CT2750	Implantable cardiac electrical stimulators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	09/12/2022	09/12/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	09/12/2022	09/12/2022
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	09/12/2022	09/12/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	09/12/2022	09/12/2022
					CT2985	Active	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	09/12/2022	09/12/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	09/12/2022	09/12/2022
					CT3001	Implantable metal	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	09/12/2022	09/12/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	09/12/2022	09/12/2022
					CT3002	Implantable synthetic polymer	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	09/12/2022	09/12/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	09/12/2022	09/12/2022
					CT3006	Implantable inorganic material	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	09/12/2022	09/12/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	09/12/2022	09/12/2022
					CT314	Electrically powered	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	09/12/2022	09/12/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	09/12/2022	09/12/2022
					CT321	Long-term surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	09/12/2022	09/12/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	09/12/2022	09/12/2022
					CT449	Generators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	09/12/2022	09/12/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	09/12/2022	09/12/2022
					CT981	Single use	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	09/12/2022	09/12/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	09/12/2022	09/12/2022
					CT983	Surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	09/12/2022	09/12/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	09/12/2022	09/12/2022
12788	2023		47265	Dual-chamber implantable pacemaker, rate-responsive	CT141	Electrodes/Leads and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem		
							F1901	Additional Surgery	A24	No Apparent Device Problem		
					CT145	Cardiovascular devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem		
							F1901	Additional Surgery	A24	No Apparent Device Problem		
					CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem		
							F1901	Additional Surgery	A24	No Apparent Device Problem		
					CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem		
							F1901	Additional Surgery	A24	No Apparent Device Problem		
					CT233	Surgical	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem		
							F1901	Additional Surgery	A24	No Apparent Device Problem		
					CT2406	Implantable	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem		
							F1901	Additional Surgery	A24	No Apparent Device Problem		
					CT2407	Active implantable	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem		
							F1901	Additional Surgery	A24	No Apparent Device Problem		
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem		
							F1901	Additional Surgery	A24	No Apparent Device Problem		
					CT2727	Active implant devices/ancillaries	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem		
							F1901	Additional Surgery	A24	No Apparent Device Problem		
					CT2750	Implantable cardiac electrical stimulators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem		
							F1901	Additional Surgery	A24	No Apparent Device Problem		
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem		
							F1901	Additional Surgery	A24	No Apparent Device Problem		
					CT2985	Active	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem		
							F1901	Additional Surgery	A24	No Apparent Device Problem		
					CT3001	Implantable metal	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem		
							F1901	Additional Surgery	A24	No Apparent Device Problem		
					CT3002	Implantable synthetic polymer	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem		
							F1901	Additional Surgery	A24	No Apparent Device Problem		
					CT3006	Implantable inorganic material	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem		
							F1901	Additional Surgery	A24	No Apparent Device Problem		
					CT314	Electrically powered	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem		
							F1901	Additional Surgery	A24	No Apparent Device Problem		
					CT321	Long-term surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem		
							F1901	Additional Surgery	A24	No Apparent Device Problem		
					CT449	Generators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem		
							F1901	Additional Surgery	A24	No Apparent Device Problem		
					CT981	Single use	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem		
							F1901	Additional Surgery	A24	No Apparent Device Problem		
					CT983	Surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem		
							F1901	Additional Surgery	A24	No Apparent Device Problem		
12789	2023	All other reportable incidents	47265	Dual-chamber implantable pacemaker, rate-responsive	CT141	Electrodes/Leads and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070909	Oversensing		
									A090801	Signal Artifact/Noise		
							F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A070909	Oversensing		
									A090801	Signal Artifact/Noise		
					CT145	Cardiovascular devices	E2403	No Clinical Signs, Symptoms or Conditions	A070909	Oversensing		

									A090801	Signal Artifact/Noise		
							F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A070909	Oversensing		
									A090801	Signal Artifact/Noise		
					CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070909	Oversensing		
									A090801	Signal Artifact/Noise		
							F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A070909	Oversensing		
									A090801	Signal Artifact/Noise		
					CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A070909	Oversensing		
									A090801	Signal Artifact/Noise		
							F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A070909	Oversensing		
									A090801	Signal Artifact/Noise		
					CT233	Surgical	E2403	No Clinical Signs, Symptoms or Conditions	A070909	Oversensing		
									A090801	Signal Artifact/Noise		
							F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A070909	Oversensing		
									A090801	Signal Artifact/Noise		
					CT2406	Implantable	E2403	No Clinical Signs, Symptoms or Conditions	A070909	Oversensing		
									A090801	Signal Artifact/Noise		
							F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A070909	Oversensing		
									A090801	Signal Artifact/Noise		
					CT2407	Active implantable	E2403	No Clinical Signs, Symptoms or Conditions	A070909	Oversensing		
									A090801	Signal Artifact/Noise		
							F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A070909	Oversensing		
									A090801	Signal Artifact/Noise		
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A070909	Oversensing		
									A090801	Signal Artifact/Noise		
							F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A070909	Oversensing		
									A090801	Signal Artifact/Noise		
					CT2727	Active implant devices/ancillaries	E2403	No Clinical Signs, Symptoms or Conditions	A070909	Oversensing		
									A090801	Signal Artifact/Noise		
							F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A070909	Oversensing		
									A090801	Signal Artifact/Noise		
					CT2750	Implantable cardiac electrical stimulators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070909	Oversensing		
									A090801	Signal Artifact/Noise		
							F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A070909	Oversensing		
									A090801	Signal Artifact/Noise		
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A070909	Oversensing		
									A090801	Signal Artifact/Noise		
							F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A070909	Oversensing		
									A090801	Signal Artifact/Noise		
					CT2985	Active	E2403	No Clinical Signs, Symptoms or Conditions	A070909	Oversensing		
									A090801	Signal Artifact/Noise		
							F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A070909	Oversensing		
									A090801	Signal Artifact/Noise		
					CT3001	Implantable metal	E2403	No Clinical Signs, Symptoms or Conditions	A070909	Oversensing		
									A090801	Signal Artifact/Noise		
							F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A070909	Oversensing		
									A090801	Signal Artifact/Noise		
					CT3002	Implantable synthetic polymer	E2403	No Clinical Signs, Symptoms or Conditions	A070909	Oversensing		
									A090801	Signal Artifact/Noise		
							F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A070909	Oversensing		
									A090801	Signal Artifact/Noise		
					CT3006	Implantable inorganic material	E2403	No Clinical Signs, Symptoms or Conditions	A070909	Oversensing		
									A090801	Signal Artifact/Noise		
							F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A070909	Oversensing		
									A090801	Signal Artifact/Noise		
					CT314	Electrically powered	E2403	No Clinical Signs, Symptoms or Conditions	A070909	Oversensing		
									A090801	Signal Artifact/Noise		
							F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A070909	Oversensing		
									A090801	Signal Artifact/Noise		
					CT321	Long-term surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A070909	Oversensing		
									A090801	Signal Artifact/Noise		
							F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A070909	Oversensing		
									A090801	Signal Artifact/Noise		
					CT449	Generators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070909	Oversensing		
									A090801	Signal Artifact/Noise		
							F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A070909	Oversensing		
									A090801	Signal Artifact/Noise		
					CT981	Single use	E2403	No Clinical Signs, Symptoms or Conditions	A070909	Oversensing		
									A090801	Signal Artifact/Noise		
							F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A070909	Oversensing		
									A090801	Signal Artifact/Noise		
					CT983	Surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A070909	Oversensing		
									A090801	Signal Artifact/Noise		
							F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A070909	Oversensing		
									A090801	Signal Artifact/Noise		
12790	2023	All other reportable incidents	17882	Physiologic-monitoring defibrillation system	CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070803	Failure to Power Up		
									A1603	Failure of Device to Self-Test		
							F27	Problem Identified During Non-Clinical Procedure	A070803	Failure to Power Up		
									A1603	Failure of Device to Self-Test		
12791	2023	All other reportable incidents	35223	Endocardial/interventricular septal pacing lead	CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070909	Oversensing		
							F26	No Health Consequences or Impact	A070909	Oversensing		
12792	2023	All other reportable incidents	47265	Dual-chamber implantable pacemaker, rate-responsive	CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A0705	Battery Problem		

									A070504	Premature Discharge of Battery		
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
12793	2023		47265	Dual-chamber implantable pacemaker, rate-responsive	CT141	Electrodes/Leads and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	31/03/2023	31/03/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	31/03/2023	31/03/2023
					CT145	Cardiovascular devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	31/03/2023	31/03/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	31/03/2023	31/03/2023
					CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	31/03/2023	31/03/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	31/03/2023	31/03/2023
					CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	31/03/2023	31/03/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	31/03/2023	31/03/2023
					CT233	Surgical	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	31/03/2023	31/03/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	31/03/2023	31/03/2023
					CT2406	Implantable	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	31/03/2023	31/03/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	31/03/2023	31/03/2023
					CT2407	Active implantable	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	31/03/2023	31/03/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	31/03/2023	31/03/2023
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	31/03/2023	31/03/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	31/03/2023	31/03/2023
					CT2727	Active implant devices/ancillaries	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	31/03/2023	31/03/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	31/03/2023	31/03/2023
					CT2750	Implantable cardiac electrical stimulators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	31/03/2023	31/03/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	31/03/2023	31/03/2023
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	31/03/2023	31/03/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	31/03/2023	31/03/2023
					CT2985	Active	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	31/03/2023	31/03/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	31/03/2023	31/03/2023
					CT3001	Implantable metal	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	31/03/2023	31/03/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	31/03/2023	31/03/2023
					CT3002	Implantable synthetic polymer	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	31/03/2023	31/03/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	31/03/2023	31/03/2023
					CT3006	Implantable inorganic material	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	31/03/2023	31/03/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	31/03/2023	31/03/2023
					CT314	Electrically powered	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	31/03/2023	31/03/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	31/03/2023	31/03/2023
					CT321	Long-term surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	31/03/2023	31/03/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	31/03/2023	31/03/2023
					CT449	Generators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	31/03/2023	31/03/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	31/03/2023	31/03/2023
					CT981	Single use	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	31/03/2023	31/03/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	31/03/2023	31/03/2023
					CT983	Surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	31/03/2023	31/03/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	31/03/2023	31/03/2023
12794	2023		47265	Dual-chamber implantable pacemaker, rate-responsive	CT141	Electrodes/Leads and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	28/03/2023	28/03/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	28/03/2023	28/03/2023
					CT145	Cardiovascular devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	28/03/2023	28/03/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	28/03/2023	28/03/2023
					CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	28/03/2023	28/03/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	28/03/2023	28/03/2023
					CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	28/03/2023	28/03/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	28/03/2023	28/03/2023
					CT233	Surgical	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	28/03/2023	28/03/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	28/03/2023	28/03/2023
					CT2406	Implantable	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	28/03/2023	28/03/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	28/03/2023	28/03/2023
					CT2407	Active implantable	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	28/03/2023	28/03/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	28/03/2023	28/03/2023
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	28/03/2023	28/03/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	28/03/2023	28/03/2023
					CT2727	Active implant devices/ancillaries	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	28/03/2023	28/03/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	28/03/2023	28/03/2023
					CT2750	Implantable cardiac electrical stimulators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	28/03/2023	28/03/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	28/03/2023	28/03/2023
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	28/03/2023	28/03/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	28/03/2023	28/03/2023
					CT2985	Active	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	28/03/2023	28/03/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	28/03/2023	28/03/2023
					CT3001	Implantable metal	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	28/03/2023	28/03/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	28/03/2023	28/03/2023
					CT3002	Implantable synthetic polymer	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	28/03/2023	28/03/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	28/03/2023	28/03/2023
					CT3006	Implantable inorganic material	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	28/03/2023	28/03/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	28/03/2023	28/03/2023
					CT314	Electrically powered	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	28/03/2023	28/03/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	28/03/2023	28/03/2023
					CT321	Long-term surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	28/03/2023	28/03/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	28/03/2023	28/03/2023
					CT449	Generators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	28/03/2023	28/03/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	28/03/2023	28/03/2023
					CT981	Single use	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	28/03/2023	28/03/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	28/03/2023	28/03/2023
					CT983	Surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	28/03/2023	28/03/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	28/03/2023	28/03/2023
12795	2023	All other reportable incidents	35852	Single-chamber implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E060109	Tachycardia	A07	Electrical /Electronic Property Problem		
									A071301	Failure to Deliver Shock/Stimulation		
							F08	Hospitalization or Prolonged Hospitalization	A07	Electrical /Electronic Property Problem		

								A071301	Failure to Deliver Shock/Stimulation			
						F1203	Life Threatening Illness or Injury	A07	Electrical /Electronic Property Problem			
								A071301	Failure to Deliver Shock/Stimulation			
						F23	Unexpected Medical Intervention	A07	Electrical /Electronic Property Problem			
								A071301	Failure to Deliver Shock/Stimulation			
12796	2023	All other reportable incidents	37265	Dual-chamber implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E2104	Electric Shock	A0712	Pacing Problem		
									A071302	Inappropriate/Inadequate Shock/Stimulation		
									A1107	Data Problem		
							F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A0712	Pacing Problem		
									A071302	Inappropriate/Inadequate Shock/Stimulation		
									A1107	Data Problem		
12797	2023	All other reportable incidents	17882	Physiologic-monitoring defibrillation system	CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A0719	Unexpected Shutdown		
							F27	Problem Identified During Non-Clinical Procedure	A0719	Unexpected Shutdown		
12798	2023	All other reportable incidents	47265	Dual-chamber implantable pacemaker, rate-responsive	CT141	Electrodes/Leads and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	09/02/2023	09/02/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	09/02/2023	09/02/2023
					CT145	Cardiovascular devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	09/02/2023	09/02/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	09/02/2023	09/02/2023
					CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	09/02/2023	09/02/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	09/02/2023	09/02/2023
					CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	09/02/2023	09/02/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	09/02/2023	09/02/2023
					CT233	Surgical	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	09/02/2023	09/02/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	09/02/2023	09/02/2023
					CT2406	Implantable	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	09/02/2023	09/02/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	09/02/2023	09/02/2023
					CT2407	Active implantable	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	09/02/2023	09/02/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	09/02/2023	09/02/2023
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	09/02/2023	09/02/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	09/02/2023	09/02/2023
					CT2727	Active implant devices/ancillaries	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	09/02/2023	09/02/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	09/02/2023	09/02/2023
					CT2750	Implantable cardiac electrical stimulators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	09/02/2023	09/02/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	09/02/2023	09/02/2023
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	09/02/2023	09/02/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	09/02/2023	09/02/2023
					CT2985	Active	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	09/02/2023	09/02/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	09/02/2023	09/02/2023
					CT3001	Implantable metal	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	09/02/2023	09/02/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	09/02/2023	09/02/2023
					CT3002	Implantable synthetic polymer	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	09/02/2023	09/02/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	09/02/2023	09/02/2023
					CT3006	Implantable inorganic material	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	09/02/2023	09/02/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	09/02/2023	09/02/2023
					CT314	Electrically powered	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	09/02/2023	09/02/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	09/02/2023	09/02/2023
					CT321	Long-term surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	09/02/2023	09/02/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	09/02/2023	09/02/2023
					CT449	Generators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	09/02/2023	09/02/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	09/02/2023	09/02/2023
					CT981	Single use	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	09/02/2023	09/02/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	09/02/2023	09/02/2023
					CT983	Surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	09/02/2023	09/02/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	09/02/2023	09/02/2023
12799	2023	All other reportable incidents	35223	Endocardial/interventricular septal pacing lead	CT1684	Cardiac pacemakers and associated devices	E0112	Dizziness	A0401	Break		
									A070909	Oversensing		
									A0722	Impedance Problem		
							E0601	Arrhythmia	A0401	Break		
									A070909	Oversensing		
									A0722	Impedance Problem		
							E1720	Skin Inflammation/ Irritation	A0401	Break		
									A070909	Oversensing		
									A0722	Impedance Problem		
							F11	Non-Serious Injury/ Illness/ Impairment	A0401	Break		
									A070909	Oversensing		
									A0722	Impedance Problem		
							F1905	Device Revision or Replacement	A0401	Break		
									A070909	Oversensing		
									A0722	Impedance Problem		
12800	2023		47265	Dual-chamber implantable pacemaker, rate-responsive	CT141	Electrodes/Leads and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	04/01/2023	04/01/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	04/01/2023	04/01/2023
					CT145	Cardiovascular devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	04/01/2023	04/01/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	04/01/2023	04/01/2023
					CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	04/01/2023	04/01/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	04/01/2023	04/01/2023
					CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	04/01/2023	04/01/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	04/01/2023	04/01/2023
					CT233	Surgical	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	04/01/2023	04/01/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	04/01/2023	04/01/2023
					CT2406	Implantable	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	04/01/2023	04/01/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	04/01/2023	04/01/2023
					CT2407	Active implantable	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	04/01/2023	04/01/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	04/01/2023	04/01/2023
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	04/01/2023	04/01/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	04/01/2023	04/01/2023
					CT2727	Active implant devices/ancillaries	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	04/01/2023	04/01/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	04/01/2023	04/01/2023
					CT2750	Implantable cardiac electrical stimulators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	04/01/2023	04/01/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	04/01/2023	04/01/2023

					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	04/01/2023	04/01/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	04/01/2023	04/01/2023
					CT2985	Active	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	04/01/2023	04/01/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	04/01/2023	04/01/2023
					CT3001	Implantable metal	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	04/01/2023	04/01/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	04/01/2023	04/01/2023
					CT3002	Implantable synthetic polymer	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	04/01/2023	04/01/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	04/01/2023	04/01/2023
					CT3006	Implantable inorganic material	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	04/01/2023	04/01/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	04/01/2023	04/01/2023
					CT314	Electrically powered	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	04/01/2023	04/01/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	04/01/2023	04/01/2023
					CT321	Long-term surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	04/01/2023	04/01/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	04/01/2023	04/01/2023
					CT449	Generators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	04/01/2023	04/01/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	04/01/2023	04/01/2023
					CT981	Single use	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	04/01/2023	04/01/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	04/01/2023	04/01/2023
					CT983	Surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	04/01/2023	04/01/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	04/01/2023	04/01/2023
12801	2023	All other reportable incidents	35223	Endocardial/interventricular septal pacing lead	CT1684	Cardiac pacemakers and associated devices	E0112	Dizziness	A0401	Break		
									A07	Electrical /Electronic Property Problem		
									A070909	Oversensing		
									A072202	Low Impedance		
							E0601	Arrhythmia	A0401	Break		
									A07	Electrical /Electronic Property Problem		
									A070909	Oversensing		
									A072202	Low Impedance		
							E1720	Skin Inflammation/ Irritation	A0401	Break		
									A07	Electrical /Electronic Property Problem		
									A070909	Oversensing		
									A072202	Low Impedance		
							F11	Non-Serious Injury/ Illness/ Impairment	A0401	Break		
									A07	Electrical /Electronic Property Problem		
									A070909	Oversensing		
									A072202	Low Impedance		
							F1905	Device Revision or Replacement	A0401	Break		
									A07	Electrical /Electronic Property Problem		
									A070909	Oversensing		
									A072202	Low Impedance		
12802	2023		47265	Dual-chamber implantable pacemaker, rate-responsive	CT141	Electrodes/Leads and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem		
							F1901	Additional Surgery	A24	No Apparent Device Problem		
					CT145	Cardiovascular devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem		
							F1901	Additional Surgery	A24	No Apparent Device Problem		
					CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem		
							F1901	Additional Surgery	A24	No Apparent Device Problem		
					CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem		
							F1901	Additional Surgery	A24	No Apparent Device Problem		
					CT233	Surgical	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem		
							F1901	Additional Surgery	A24	No Apparent Device Problem		
					CT2406	Implantable	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem		
							F1901	Additional Surgery	A24	No Apparent Device Problem		
					CT2407	Active implantable	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem		
							F1901	Additional Surgery	A24	No Apparent Device Problem		
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem		
							F1901	Additional Surgery	A24	No Apparent Device Problem		
					CT2727	Active implant devices/ancillaries	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem		
							F1901	Additional Surgery	A24	No Apparent Device Problem		
					CT2750	Implantable cardiac electrical stimulators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem		
							F1901	Additional Surgery	A24	No Apparent Device Problem		
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem		
							F1901	Additional Surgery	A24	No Apparent Device Problem		
					CT2985	Active	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem		
							F1901	Additional Surgery	A24	No Apparent Device Problem		
					CT3001	Implantable metal	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem		
							F1901	Additional Surgery	A24	No Apparent Device Problem		
					CT3002	Implantable synthetic polymer	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem		
							F1901	Additional Surgery	A24	No Apparent Device Problem		
					CT3006	Implantable inorganic material	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem		
							F1901	Additional Surgery	A24	No Apparent Device Problem		
					CT314	Electrically powered	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem		
							F1901	Additional Surgery	A24	No Apparent Device Problem		
					CT321	Long-term surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem		
							F1901	Additional Surgery	A24	No Apparent Device Problem		
					CT449	Generators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem		
							F1901	Additional Surgery	A24	No Apparent Device Problem		
					CT981	Single use	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem		
							F1901	Additional Surgery	A24	No Apparent Device Problem		
					CT983	Surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem		
							F1901	Additional Surgery	A24	No Apparent Device Problem		
12803	2023	All other reportable incidents	47265	Dual-chamber implantable pacemaker, rate-responsive	CT141	Electrodes/Leads and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A072201	High Impedance		
							F26	No Health Consequences or Impact	A072201	High Impedance		
					CT145	Cardiovascular devices	E2403	No Clinical Signs, Symptoms or Conditions	A072201	High Impedance		
							F26	No Health Consequences or Impact	A072201	High Impedance		
					CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A072201	High Impedance		
							F26	No Health Consequences or Impact	A072201	High Impedance		
					CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A072201	High Impedance		
							F26	No Health Consequences or Impact	A072201	High Impedance		

					CT233	Surgical	E2403	No Clinical Signs, Symptoms or Conditions	A072201	High Impedance		
							F26	No Health Consequences or Impact	A072201	High Impedance		
					CT2406	Implantable	E2403	No Clinical Signs, Symptoms or Conditions	A072201	High Impedance		
							F26	No Health Consequences or Impact	A072201	High Impedance		
					CT2407	Active implantable	E2403	No Clinical Signs, Symptoms or Conditions	A072201	High Impedance		
							F26	No Health Consequences or Impact	A072201	High Impedance		
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A072201	High Impedance		
							F26	No Health Consequences or Impact	A072201	High Impedance		
					CT2727	Active implant devices/ancillaries	E2403	No Clinical Signs, Symptoms or Conditions	A072201	High Impedance		
							F26	No Health Consequences or Impact	A072201	High Impedance		
					CT2750	Implantable cardiac electrical stimulators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A072201	High Impedance		
							F26	No Health Consequences or Impact	A072201	High Impedance		
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A072201	High Impedance		
							F26	No Health Consequences or Impact	A072201	High Impedance		
					CT2985	Active	E2403	No Clinical Signs, Symptoms or Conditions	A072201	High Impedance		
							F26	No Health Consequences or Impact	A072201	High Impedance		
					CT3001	Implantable metal	E2403	No Clinical Signs, Symptoms or Conditions	A072201	High Impedance		
							F26	No Health Consequences or Impact	A072201	High Impedance		
					CT3002	Implantable synthetic polymer	E2403	No Clinical Signs, Symptoms or Conditions	A072201	High Impedance		
							F26	No Health Consequences or Impact	A072201	High Impedance		
					CT3006	Implantable inorganic material	E2403	No Clinical Signs, Symptoms or Conditions	A072201	High Impedance		
							F26	No Health Consequences or Impact	A072201	High Impedance		
					CT314	Electrically powered	E2403	No Clinical Signs, Symptoms or Conditions	A072201	High Impedance		
							F26	No Health Consequences or Impact	A072201	High Impedance		
					CT321	Long-term surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A072201	High Impedance		
							F26	No Health Consequences or Impact	A072201	High Impedance		
					CT449	Generators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A072201	High Impedance		
							F26	No Health Consequences or Impact	A072201	High Impedance		
					CT981	Single use	E2403	No Clinical Signs, Symptoms or Conditions	A072201	High Impedance		
							F26	No Health Consequences or Impact	A072201	High Impedance		
					CT983	Surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A072201	High Impedance		
							F26	No Health Consequences or Impact	A072201	High Impedance		
12804	2023	All other reportable incidents	36241	Pacemaker/defibrillator lead	CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A072201	High Impedance		
							F26	No Health Consequences or Impact	A072201	High Impedance		
12805	2023	All other reportable incidents	35852	Single-chamber implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E2104	Electric Shock	A070909	Oversensing		
									A070910	Undersensing		
									A071302	Inappropriate/Inadequate Shock/Stimulation		
									A090801	Signal Artifact/Noise		
							F08	Hospitalization or Prolonged Hospitalization	A070909	Oversensing		
									A070910	Undersensing		
									A071302	Inappropriate/Inadequate Shock/Stimulation		
									A090801	Signal Artifact/Noise		
							F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A070909	Oversensing		
									A070910	Undersensing		
									A071302	Inappropriate/Inadequate Shock/Stimulation		
									A090801	Signal Artifact/Noise		
							F1905	Device Revision or Replacement	A070909	Oversensing		
									A070910	Undersensing		
									A071302	Inappropriate/Inadequate Shock/Stimulation		
									A090801	Signal Artifact/Noise		
							F2203	Imaging Required	A070909	Oversensing		
									A070910	Undersensing		
									A071302	Inappropriate/Inadequate Shock/Stimulation		
									A090801	Signal Artifact/Noise		
12806	2023	All other reportable incidents	47265	Dual-chamber implantable pacemaker, rate-responsive	CT141	Electrodes/Leads and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	22/03/2023	22/03/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	22/03/2023	22/03/2023
					CT145	Cardiovascular devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	22/03/2023	22/03/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	22/03/2023	22/03/2023
					CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	22/03/2023	22/03/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	22/03/2023	22/03/2023
					CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	22/03/2023	22/03/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	22/03/2023	22/03/2023
					CT233	Surgical	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	22/03/2023	22/03/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	22/03/2023	22/03/2023
					CT2406	Implantable	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	22/03/2023	22/03/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	22/03/2023	22/03/2023
					CT2407	Active implantable	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	22/03/2023	22/03/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	22/03/2023	22/03/2023
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	22/03/2023	22/03/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	22/03/2023	22/03/2023
					CT2727	Active implant devices/ancillaries	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	22/03/2023	22/03/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	22/03/2023	22/03/2023
					CT2750	Implantable cardiac electrical stimulators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	22/03/2023	22/03/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	22/03/2023	22/03/2023
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	22/03/2023	22/03/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	22/03/2023	22/03/2023
					CT2985	Active	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	22/03/2023	22/03/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	22/03/2023	22/03/2023
					CT3001	Implantable metal	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	22/03/2023	22/03/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	22/03/2023	22/03/2023
					CT3002	Implantable synthetic polymer	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	22/03/2023	22/03/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	22/03/2023	22/03/2023
					CT3006	Implantable inorganic material	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	22/03/2023	22/03/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	22/03/2023	22/03/2023
					CT314	Electrically powered	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	22/03/2023	22/03/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	22/03/2023	22/03/2023
					CT321	Long-term surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	22/03/2023	22/03/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	22/03/2023	22/03/2023

					CT449	Generators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	22/03/2023	22/03/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	22/03/2023	22/03/2023
					CT981	Single use	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	22/03/2023	22/03/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	22/03/2023	22/03/2023
					CT983	Surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	22/03/2023	22/03/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	22/03/2023	22/03/2023
12807	2023		47265	Dual-chamber implantable pacemaker, rate-responsive	CT141	Electrodes/Leads and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem		
							F1901	Additional Surgery	A24	No Apparent Device Problem		
					CT145	Cardiovascular devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem		
							F1901	Additional Surgery	A24	No Apparent Device Problem		
					CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem		
							F1901	Additional Surgery	A24	No Apparent Device Problem		
					CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem		
							F1901	Additional Surgery	A24	No Apparent Device Problem		
					CT233	Surgical	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem		
							F1901	Additional Surgery	A24	No Apparent Device Problem		
					CT2406	Implantable	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem		
							F1901	Additional Surgery	A24	No Apparent Device Problem		
					CT2407	Active implantable	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem		
							F1901	Additional Surgery	A24	No Apparent Device Problem		
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem		
							F1901	Additional Surgery	A24	No Apparent Device Problem		
					CT2727	Active implant devices/ancillaries	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem		
							F1901	Additional Surgery	A24	No Apparent Device Problem		
					CT2750	Implantable cardiac electrical stimulators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem		
							F1901	Additional Surgery	A24	No Apparent Device Problem		
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem		
							F1901	Additional Surgery	A24	No Apparent Device Problem		
					CT2985	Active	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem		
							F1901	Additional Surgery	A24	No Apparent Device Problem		
					CT3001	Implantable metal	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem		
							F1901	Additional Surgery	A24	No Apparent Device Problem		
					CT3002	Implantable synthetic polymer	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem		
							F1901	Additional Surgery	A24	No Apparent Device Problem		
					CT3006	Implantable inorganic material	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem		
							F1901	Additional Surgery	A24	No Apparent Device Problem		
					CT314	Electrically powered	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem		
							F1901	Additional Surgery	A24	No Apparent Device Problem		
					CT321	Long-term surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem		
							F1901	Additional Surgery	A24	No Apparent Device Problem		
					CT449	Generators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem		
							F1901	Additional Surgery	A24	No Apparent Device Problem		
					CT981	Single use	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem		
							F1901	Additional Surgery	A24	No Apparent Device Problem		
					CT983	Surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem		
							F1901	Additional Surgery	A24	No Apparent Device Problem		
12808	2023		47265	Dual-chamber implantable pacemaker, rate-responsive	CT141	Electrodes/Leads and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	04/01/2023	04/01/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	04/01/2023	04/01/2023
					CT145	Cardiovascular devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	04/01/2023	04/01/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	04/01/2023	04/01/2023
					CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	04/01/2023	04/01/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	04/01/2023	04/01/2023
					CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	04/01/2023	04/01/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	04/01/2023	04/01/2023
					CT233	Surgical	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	04/01/2023	04/01/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	04/01/2023	04/01/2023
					CT2406	Implantable	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	04/01/2023	04/01/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	04/01/2023	04/01/2023
					CT2407	Active implantable	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	04/01/2023	04/01/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	04/01/2023	04/01/2023
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	04/01/2023	04/01/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	04/01/2023	04/01/2023
					CT2727	Active implant devices/ancillaries	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	04/01/2023	04/01/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	04/01/2023	04/01/2023
					CT2750	Implantable cardiac electrical stimulators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	04/01/2023	04/01/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	04/01/2023	04/01/2023
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	04/01/2023	04/01/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	04/01/2023	04/01/2023
					CT2985	Active	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	04/01/2023	04/01/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	04/01/2023	04/01/2023
					CT3001	Implantable metal	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	04/01/2023	04/01/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	04/01/2023	04/01/2023
					CT3002	Implantable synthetic polymer	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	04/01/2023	04/01/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	04/01/2023	04/01/2023
					CT3006	Implantable inorganic material	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	04/01/2023	04/01/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	04/01/2023	04/01/2023
					CT314	Electrically powered	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	04/01/2023	04/01/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	04/01/2023	04/01/2023
					CT321	Long-term surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	04/01/2023	04/01/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	04/01/2023	04/01/2023
					CT449	Generators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	04/01/2023	04/01/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	04/01/2023	04/01/2023
					CT981	Single use	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	04/01/2023	04/01/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	04/01/2023	04/01/2023
					CT983	Surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	04/01/2023	04/01/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	04/01/2023	04/01/2023
12809	2023	All other reportable incidents	35852	Single-chamber implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem	27/03/2023	27/03/2023
									A070504	Premature Discharge of Battery	27/03/2023	27/03/2023

							F08	Hospitalization or Prolonged Hospitalization	A0705	Battery Problem	27/03/2023	27/03/2023
									A070504	Premature Discharge of Battery	27/03/2023	27/03/2023
							F1905	Device Revision or Replacement	A0705	Battery Problem	27/03/2023	27/03/2023
									A070504	Premature Discharge of Battery	27/03/2023	27/03/2023
12810	2023	All other reportable incidents	17882	Physiologic-monitoring defibrillation system	CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070803	Failure to Power Up		
							F27	Problem Identified During Non-Clinical Procedure	A070803	Failure to Power Up		
12811	2023	All other reportable incidents	60789	Intracardiac pacemaker	CT130	Catheters and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A0504	Leak/Splash		
							F26	No Health Consequences or Impact	A0504	Leak/Splash		
					CT145	Cardiovascular devices	E2403	No Clinical Signs, Symptoms or Conditions	A0504	Leak/Splash		
							F26	No Health Consequences or Impact	A0504	Leak/Splash		
					CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A0504	Leak/Splash		
							F26	No Health Consequences or Impact	A0504	Leak/Splash		
					CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A0504	Leak/Splash		
							F26	No Health Consequences or Impact	A0504	Leak/Splash		
					CT233	Surgical	E2403	No Clinical Signs, Symptoms or Conditions	A0504	Leak/Splash		
							F26	No Health Consequences or Impact	A0504	Leak/Splash		
					CT2406	Implantable	E2403	No Clinical Signs, Symptoms or Conditions	A0504	Leak/Splash		
							F26	No Health Consequences or Impact	A0504	Leak/Splash		
					CT2407	Active implantable	E2403	No Clinical Signs, Symptoms or Conditions	A0504	Leak/Splash		
							F26	No Health Consequences or Impact	A0504	Leak/Splash		
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A0504	Leak/Splash		
							F26	No Health Consequences or Impact	A0504	Leak/Splash		
					CT2727	Active implant devices/ancillaries	E2403	No Clinical Signs, Symptoms or Conditions	A0504	Leak/Splash		
							F26	No Health Consequences or Impact	A0504	Leak/Splash		
					CT2750	Implantable cardiac electrical stimulators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A0504	Leak/Splash		
							F26	No Health Consequences or Impact	A0504	Leak/Splash		
					CT2818	Catheterization/Cannulation/Fluid path devices	E2403	No Clinical Signs, Symptoms or Conditions	A0504	Leak/Splash		
							F26	No Health Consequences or Impact	A0504	Leak/Splash		
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A0504	Leak/Splash		
							F26	No Health Consequences or Impact	A0504	Leak/Splash		
					CT2985	Active	E2403	No Clinical Signs, Symptoms or Conditions	A0504	Leak/Splash		
							F26	No Health Consequences or Impact	A0504	Leak/Splash		
					CT3001	Implantable metal	E2403	No Clinical Signs, Symptoms or Conditions	A0504	Leak/Splash		
							F26	No Health Consequences or Impact	A0504	Leak/Splash		
					CT3002	Implantable synthetic polymer	E2403	No Clinical Signs, Symptoms or Conditions	A0504	Leak/Splash		
							F26	No Health Consequences or Impact	A0504	Leak/Splash		
					CT3006	Implantable inorganic material	E2403	No Clinical Signs, Symptoms or Conditions	A0504	Leak/Splash		
							F26	No Health Consequences or Impact	A0504	Leak/Splash		
					CT314	Electrically powered	E2403	No Clinical Signs, Symptoms or Conditions	A0504	Leak/Splash		
							F26	No Health Consequences or Impact	A0504	Leak/Splash		
					CT321	Long-term surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A0504	Leak/Splash		
							F26	No Health Consequences or Impact	A0504	Leak/Splash		
					CT981	Single use	E2403	No Clinical Signs, Symptoms or Conditions	A0504	Leak/Splash		
							F26	No Health Consequences or Impact	A0504	Leak/Splash		
					CT983	Surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A0504	Leak/Splash		
							F26	No Health Consequences or Impact	A0504	Leak/Splash		
12812	2023	All other reportable incidents	17882	Physiologic-monitoring defibrillation system	CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A090811	Unable to Obtain Readings		
							F27	Problem Identified During Non-Clinical Procedure	A090811	Unable to Obtain Readings		
12813	2023	All other reportable incidents	17882	Physiologic-monitoring defibrillation system	CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem		
							F27	Problem Identified During Non-Clinical Procedure	A0705	Battery Problem		
12814	2023	All other reportable incidents	35223	Endocardial/interventricular septal pacing lead	CT1684	Cardiac pacemakers and associated devices	E0112	Dizziness	A040101	Fracture		
									A070909	Oversensing		
							E0601	Arrhythmia	A040101	Fracture		
									A070909	Oversensing		
							F1901	Additional Surgery	A040101	Fracture		
									A070909	Oversensing		
12815	2023	All other reportable incidents	35223	Endocardial/interventricular septal pacing lead	CT1684	Cardiac pacemakers and associated devices	E060109	Tachycardia	A070909	Oversensing		
							F1001	Absence of Treatment	A070909	Oversensing		
							F1203	Life Threatening Illness or Injury	A070909	Oversensing		
12816	2023		47270	Cardiac resynchronization therapy implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	06/04/2023	06/04/2023
							F1901	Additional Surgery	A070504	Premature Discharge of Battery	06/04/2023	06/04/2023
12817	2023		47265	Dual-chamber implantable pacemaker, rate-responsive	CT141	Electrodes/Leads and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem		
							F1901	Additional Surgery	A24	No Apparent Device Problem		
					CT145	Cardiovascular devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem		
							F1901	Additional Surgery	A24	No Apparent Device Problem		
					CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem		
							F1901	Additional Surgery	A24	No Apparent Device Problem		
					CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem		
							F1901	Additional Surgery	A24	No Apparent Device Problem		
					CT233	Surgical	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem		
							F1901	Additional Surgery	A24	No Apparent Device Problem		
					CT2406	Implantable	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem		
							F1901	Additional Surgery	A24	No Apparent Device Problem		
					CT2407	Active implantable	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem		
							F1901	Additional Surgery	A24	No Apparent Device Problem		
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem		
							F1901	Additional Surgery	A24	No Apparent Device Problem		
					CT2727	Active implant devices/ancillaries	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem		
							F1901	Additional Surgery	A24	No Apparent Device Problem		
					CT2750	Implantable cardiac electrical stimulators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem		
							F1901	Additional Surgery	A24	No Apparent Device Problem		
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem		
							F1901	Additional Surgery	A24	No Apparent Device Problem		
					CT2985	Active	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem		
							F1901	Additional Surgery	A24	No Apparent Device Problem		
					CT3001	Implantable metal	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem		
							F1901	Additional Surgery	A24	No Apparent Device Problem		
					CT3002	Implantable synthetic polymer	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem		

						F1901	Additional Surgery	A24	No Apparent Device Problem		
				CT3006	Implantable inorganic material	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem		
						F1901	Additional Surgery	A24	No Apparent Device Problem		
				CT314	Electrically powered	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem		
						F1901	Additional Surgery	A24	No Apparent Device Problem		
				CT321	Long-term surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem		
						F1901	Additional Surgery	A24	No Apparent Device Problem		
				CT449	Generators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem		
						F1901	Additional Surgery	A24	No Apparent Device Problem		
				CT981	Single use	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem		
						F1901	Additional Surgery	A24	No Apparent Device Problem		
				CT983	Surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem		
						F1901	Additional Surgery	A24	No Apparent Device Problem		
12818	2023		47265	Dual-chamber implantable pacemaker, rate-responsive	CT141	Electrodes/Leads and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	28/02/2023	28/02/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	28/02/2023	28/02/2023
					CT145	Cardiovascular devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	28/02/2023	28/02/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	28/02/2023	28/02/2023
					CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	28/02/2023	28/02/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	28/02/2023	28/02/2023
					CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	28/02/2023	28/02/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	28/02/2023	28/02/2023
					CT233	Surgical	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	28/02/2023	28/02/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	28/02/2023	28/02/2023
					CT2406	Implantable	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	28/02/2023	28/02/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	28/02/2023	28/02/2023
					CT2407	Active implantable	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	28/02/2023	28/02/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	28/02/2023	28/02/2023
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	28/02/2023	28/02/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	28/02/2023	28/02/2023
					CT2727	Active implant devices/ancillaries	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	28/02/2023	28/02/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	28/02/2023	28/02/2023
					CT2750	Implantable cardiac electrical stimulators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	28/02/2023	28/02/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	28/02/2023	28/02/2023
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	28/02/2023	28/02/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	28/02/2023	28/02/2023
					CT2985	Active	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	28/02/2023	28/02/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	28/02/2023	28/02/2023
					CT3001	Implantable metal	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	28/02/2023	28/02/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	28/02/2023	28/02/2023
					CT3002	Implantable synthetic polymer	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	28/02/2023	28/02/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	28/02/2023	28/02/2023
					CT3006	Implantable inorganic material	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	28/02/2023	28/02/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	28/02/2023	28/02/2023
					CT314	Electrically powered	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	28/02/2023	28/02/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	28/02/2023	28/02/2023
					CT321	Long-term surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	28/02/2023	28/02/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	28/02/2023	28/02/2023
					CT449	Generators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	28/02/2023	28/02/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	28/02/2023	28/02/2023
					CT981	Single use	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	28/02/2023	28/02/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	28/02/2023	28/02/2023
					CT983	Surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	28/02/2023	28/02/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	28/02/2023	28/02/2023
12819	2023	All other reportable incidents	47263	Cardiac resynchronization therapy implantable pacemaker	CT1685	Cardiac pacemakers	E2007	Fall	A070909	Oversensing	03/04/2023	03/04/2023
									A0712	Pacing Problem	03/04/2023	03/04/2023
									A0908	Incorrect, Inadequate or Imprecise Result or Readings	03/04/2023	03/04/2023
							F08	Hospitalization or Prolonged Hospitalization	A070909	Oversensing	03/04/2023	03/04/2023
									A0712	Pacing Problem	03/04/2023	03/04/2023
									A0908	Incorrect, Inadequate or Imprecise Result or Readings	03/04/2023	03/04/2023
							F1905	Device Revision or Replacement	A070909	Oversensing	03/04/2023	03/04/2023
									A0712	Pacing Problem	03/04/2023	03/04/2023
									A0908	Incorrect, Inadequate or Imprecise Result or Readings	03/04/2023	03/04/2023
					CT257	Implantable cardiac pacemakers	E2007	Fall	A070909	Oversensing	03/04/2023	03/04/2023
									A0712	Pacing Problem	03/04/2023	03/04/2023
									A0908	Incorrect, Inadequate or Imprecise Result or Readings	03/04/2023	03/04/2023
							F08	Hospitalization or Prolonged Hospitalization	A070909	Oversensing	03/04/2023	03/04/2023
									A0712	Pacing Problem	03/04/2023	03/04/2023
									A0908	Incorrect, Inadequate or Imprecise Result or Readings	03/04/2023	03/04/2023
							F1905	Device Revision or Replacement	A070909	Oversensing	03/04/2023	03/04/2023
									A0712	Pacing Problem	03/04/2023	03/04/2023
									A0908	Incorrect, Inadequate or Imprecise Result or Readings	03/04/2023	03/04/2023
					CT2866	Implantable cardiac electrical stimulators	E2007	Fall	A070909	Oversensing	03/04/2023	03/04/2023
									A0712	Pacing Problem	03/04/2023	03/04/2023
									A0908	Incorrect, Inadequate or Imprecise Result or Readings	03/04/2023	03/04/2023
							F08	Hospitalization or Prolonged Hospitalization	A070909	Oversensing	03/04/2023	03/04/2023
									A0712	Pacing Problem	03/04/2023	03/04/2023
									A0908	Incorrect, Inadequate or Imprecise Result or Readings	03/04/2023	03/04/2023
							F1905	Device Revision or Replacement	A070909	Oversensing	03/04/2023	03/04/2023
									A0712	Pacing Problem	03/04/2023	03/04/2023
									A0908	Incorrect, Inadequate or Imprecise Result or Readings	03/04/2023	03/04/2023
12820	2023	All other reportable incidents	47265	Dual-chamber implantable pacemaker, rate-	CT141	Electrodes/Leads and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	03/01/2023	03/01/2023

				responsive			F1901	Additional Surgery	A24	No Apparent Device Problem	03/01/2023	03/01/2023
					CT145	Cardiovascular devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	03/01/2023	03/01/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	03/01/2023	03/01/2023
					CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	03/01/2023	03/01/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	03/01/2023	03/01/2023
					CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	03/01/2023	03/01/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	03/01/2023	03/01/2023
					CT233	Surgical	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	03/01/2023	03/01/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	03/01/2023	03/01/2023
					CT2406	Implantable	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	03/01/2023	03/01/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	03/01/2023	03/01/2023
					CT2407	Active implantable	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	03/01/2023	03/01/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	03/01/2023	03/01/2023
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	03/01/2023	03/01/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	03/01/2023	03/01/2023
					CT2727	Active implant devices/ancillaries	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	03/01/2023	03/01/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	03/01/2023	03/01/2023
					CT2750	Implantable cardiac electrical stimulators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	03/01/2023	03/01/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	03/01/2023	03/01/2023
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	03/01/2023	03/01/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	03/01/2023	03/01/2023
					CT2985	Active	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	03/01/2023	03/01/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	03/01/2023	03/01/2023
					CT3001	Implantable metal	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	03/01/2023	03/01/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	03/01/2023	03/01/2023
					CT3002	Implantable synthetic polymer	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	03/01/2023	03/01/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	03/01/2023	03/01/2023
					CT3006	Implantable inorganic material	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	03/01/2023	03/01/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	03/01/2023	03/01/2023
					CT314	Electrically powered	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	03/01/2023	03/01/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	03/01/2023	03/01/2023
					CT321	Long-term surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	03/01/2023	03/01/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	03/01/2023	03/01/2023
					CT449	Generators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	03/01/2023	03/01/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	03/01/2023	03/01/2023
					CT981	Single use	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	03/01/2023	03/01/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	03/01/2023	03/01/2023
					CT983	Surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	03/01/2023	03/01/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	03/01/2023	03/01/2023
12821	2023	All other reportable incidents	47265	Dual-chamber implantable pacemaker, rate-responsive	CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
12822	2023	All other reportable incidents	37265	Dual-chamber implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A072201	High Impedance		
							F26	No Health Consequences or Impact	A072201	High Impedance		
12823	2023	All other reportable incidents	36241	Pacemaker/defibrillator lead	CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A072201	High Impedance		
							F26	No Health Consequences or Impact	A072201	High Impedance		
12824	2023	All other reportable incidents	37265	Dual-chamber implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A070909	Oversensing		
									A090801	Signal Artifact/Noise		
							F26	No Health Consequences or Impact	A070909	Oversensing		
									A090801	Signal Artifact/Noise		
12825	2023	All other reportable incidents	36241	Pacemaker/defibrillator lead	CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A0203	Defective Device		
									A070909	Oversensing		
									A090801	Signal Artifact/Noise		
							F26	No Health Consequences or Impact	A0203	Defective Device		
									A070909	Oversensing		
									A090801	Signal Artifact/Noise		
12826	2023	All other reportable incidents	47267	Single-chamber implantable pacemaker, rate-responsive	CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
12827	2023	All other reportable incidents	17882	Physiologic-monitoring defibrillation system	CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070803	Failure to Power Up		
							F27	Problem Identified During Non-Clinical Procedure	A070803	Failure to Power Up		
12828	2023	All other reportable incidents	35223	Endocardial/interventricular septal pacing lead	CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070909	Oversensing		
									A070910	Undersensing		
							F26	No Health Consequences or Impact	A070909	Oversensing		
									A070910	Undersensing		
12829	2023	All other reportable incidents	35852	Single-chamber implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E2104	Electric Shock	A070909	Oversensing		
									A071302	Inappropriate/Inadequate Shock/Stimulation		
									A1301	Failure to Read Input Signal		
							F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A070909	Oversensing		
									A071302	Inappropriate/Inadequate Shock/Stimulation		
									A1301	Failure to Read Input Signal		
12830	2023	All other reportable incidents	35223	Endocardial/interventricular septal pacing lead	CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A0401	Break		
									A070909	Oversensing		
							F26	No Health Consequences or Impact	A0401	Break		
									A070909	Oversensing		
12831	2023	All other reportable incidents	47265	Dual-chamber implantable pacemaker, rate-responsive	CT141	Electrodes/Leads and associated devices	E011903	Syncope/Fainting	A070909	Oversensing	10/03/2023	10/03/2023
									A0712	Pacing Problem	10/03/2023	10/03/2023
									A072201	High Impedance	10/03/2023	10/03/2023
									A090801	Signal Artifact/Noise	10/03/2023	10/03/2023
							E060101	Asystole	A070909	Oversensing	10/03/2023	10/03/2023
									A0712	Pacing Problem	10/03/2023	10/03/2023
									A072201	High Impedance	10/03/2023	10/03/2023

								A090801	Signal Artifact/Noise	10/03/2023	10/03/2023
						E060106	Heart Block	A070909	Oversensing	10/03/2023	10/03/2023
								A0712	Pacing Problem	10/03/2023	10/03/2023
								A072201	High Impedance	10/03/2023	10/03/2023
								A090801	Signal Artifact/Noise	10/03/2023	10/03/2023
						F08	Hospitalization or Prolonged Hospitalization	A070909	Oversensing	10/03/2023	10/03/2023
								A0712	Pacing Problem	10/03/2023	10/03/2023
								A072201	High Impedance	10/03/2023	10/03/2023
								A090801	Signal Artifact/Noise	10/03/2023	10/03/2023
						F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A070909	Oversensing	10/03/2023	10/03/2023
								A0712	Pacing Problem	10/03/2023	10/03/2023
								A072201	High Impedance	10/03/2023	10/03/2023
								A090801	Signal Artifact/Noise	10/03/2023	10/03/2023
						F1203	Life Threatening Illness or Injury	A070909	Oversensing	10/03/2023	10/03/2023
								A0712	Pacing Problem	10/03/2023	10/03/2023
								A072201	High Impedance	10/03/2023	10/03/2023
								A090801	Signal Artifact/Noise	10/03/2023	10/03/2023
						F1905	Device Revision or Replacement	A070909	Oversensing	10/03/2023	10/03/2023
								A0712	Pacing Problem	10/03/2023	10/03/2023
								A072201	High Impedance	10/03/2023	10/03/2023
								A090801	Signal Artifact/Noise	10/03/2023	10/03/2023
					CT145	Cardiovascular devices	E011903	Syncope/Fainting	A070909	Oversensing	10/03/2023	10/03/2023
									A0712	Pacing Problem	10/03/2023	10/03/2023
									A072201	High Impedance	10/03/2023	10/03/2023
									A090801	Signal Artifact/Noise	10/03/2023	10/03/2023
							E060101	Asystole	A070909	Oversensing	10/03/2023	10/03/2023
									A0712	Pacing Problem	10/03/2023	10/03/2023
									A072201	High Impedance	10/03/2023	10/03/2023
									A090801	Signal Artifact/Noise	10/03/2023	10/03/2023
							E060106	Heart Block	A070909	Oversensing	10/03/2023	10/03/2023
									A0712	Pacing Problem	10/03/2023	10/03/2023
									A072201	High Impedance	10/03/2023	10/03/2023
									A090801	Signal Artifact/Noise	10/03/2023	10/03/2023
							F08	Hospitalization or Prolonged Hospitalization	A070909	Oversensing	10/03/2023	10/03/2023
									A0712	Pacing Problem	10/03/2023	10/03/2023
									A072201	High Impedance	10/03/2023	10/03/2023
									A090801	Signal Artifact/Noise	10/03/2023	10/03/2023
							F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A070909	Oversensing	10/03/2023	10/03/2023
									A0712	Pacing Problem	10/03/2023	10/03/2023
									A072201	High Impedance	10/03/2023	10/03/2023
									A090801	Signal Artifact/Noise	10/03/2023	10/03/2023
							F1203	Life Threatening Illness or Injury	A070909	Oversensing	10/03/2023	10/03/2023
									A0712	Pacing Problem	10/03/2023	10/03/2023
									A072201	High Impedance	10/03/2023	10/03/2023
									A090801	Signal Artifact/Noise	10/03/2023	10/03/2023
							F1905	Device Revision or Replacement	A070909	Oversensing	10/03/2023	10/03/2023
									A0712	Pacing Problem	10/03/2023	10/03/2023
									A072201	High Impedance	10/03/2023	10/03/2023
									A090801	Signal Artifact/Noise	10/03/2023	10/03/2023
					CT1684	Cardiac pacemakers and associated devices	E011903	Syncope/Fainting	A070909	Oversensing	10/03/2023	10/03/2023
									A0712	Pacing Problem	10/03/2023	10/03/2023
									A072201	High Impedance	10/03/2023	10/03/2023
									A090801	Signal Artifact/Noise	10/03/2023	10/03/2023
							E060101	Asystole	A070909	Oversensing	10/03/2023	10/03/2023
									A0712	Pacing Problem	10/03/2023	10/03/2023
									A072201	High Impedance	10/03/2023	10/03/2023
									A090801	Signal Artifact/Noise	10/03/2023	10/03/2023
							E060106	Heart Block	A070909	Oversensing	10/03/2023	10/03/2023
									A0712	Pacing Problem	10/03/2023	10/03/2023
									A072201	High Impedance	10/03/2023	10/03/2023
									A090801	Signal Artifact/Noise	10/03/2023	10/03/2023
							F08	Hospitalization or Prolonged Hospitalization	A070909	Oversensing	10/03/2023	10/03/2023
									A0712	Pacing Problem	10/03/2023	10/03/2023
									A072201	High Impedance	10/03/2023	10/03/2023
									A090801	Signal Artifact/Noise	10/03/2023	10/03/2023
							F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A070909	Oversensing	10/03/2023	10/03/2023
									A0712	Pacing Problem	10/03/2023	10/03/2023
									A072201	High Impedance	10/03/2023	10/03/2023
									A090801	Signal Artifact/Noise	10/03/2023	10/03/2023
							F1203	Life Threatening Illness or Injury	A070909	Oversensing	10/03/2023	10/03/2023
									A0712	Pacing Problem	10/03/2023	10/03/2023
									A072201	High Impedance	10/03/2023	10/03/2023
									A090801	Signal Artifact/Noise	10/03/2023	10/03/2023
							F1905	Device Revision or Replacement	A070909	Oversensing	10/03/2023	10/03/2023
									A0712	Pacing Problem	10/03/2023	10/03/2023
									A072201	High Impedance	10/03/2023	10/03/2023
									A090801	Signal Artifact/Noise	10/03/2023	10/03/2023
					CT1685	Cardiac pacemakers	E011903	Syncope/Fainting	A070909	Oversensing	10/03/2023	10/03/2023
									A0712	Pacing Problem	10/03/2023	10/03/2023
									A072201	High Impedance	10/03/2023	10/03/2023
									A090801	Signal Artifact/Noise	10/03/2023	10/03/2023
							E060101	Asystole	A070909	Oversensing	10/03/2023	10/03/2023
									A0712	Pacing Problem	10/03/2023	10/03/2023
									A072201	High Impedance	10/03/2023	10/03/2023
									A090801	Signal Artifact/Noise	10/03/2023	10/03/2023
							E060106	Heart Block	A070909	Oversensing	10/03/2023	10/03/2023
									A0712	Pacing Problem	10/03/2023	10/03/2023
									A072201	High Impedance	10/03/2023	10/03/2023

								A090801	Signal Artifact/Noise	10/03/2023	10/03/2023
							F08	Hospitalization or Prolonged Hospitalization	A070909	Oversensing	10/03/2023	10/03/2023
									A0712	Pacing Problem	10/03/2023	10/03/2023
									A072201	High Impedance	10/03/2023	10/03/2023
									A090801	Signal Artifact/Noise	10/03/2023	10/03/2023
							F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A070909	Oversensing	10/03/2023	10/03/2023
									A0712	Pacing Problem	10/03/2023	10/03/2023
									A072201	High Impedance	10/03/2023	10/03/2023
									A090801	Signal Artifact/Noise	10/03/2023	10/03/2023
							F1203	Life Threatening Illness or Injury	A070909	Oversensing	10/03/2023	10/03/2023
									A0712	Pacing Problem	10/03/2023	10/03/2023
									A072201	High Impedance	10/03/2023	10/03/2023
									A090801	Signal Artifact/Noise	10/03/2023	10/03/2023
							F1905	Device Revision or Replacement	A070909	Oversensing	10/03/2023	10/03/2023
									A0712	Pacing Problem	10/03/2023	10/03/2023
									A072201	High Impedance	10/03/2023	10/03/2023
									A090801	Signal Artifact/Noise	10/03/2023	10/03/2023
					CT233	Surgical	E011903	Syncope/Fainting	A070909	Oversensing	10/03/2023	10/03/2023
									A0712	Pacing Problem	10/03/2023	10/03/2023
									A072201	High Impedance	10/03/2023	10/03/2023
									A090801	Signal Artifact/Noise	10/03/2023	10/03/2023
							E060101	Asystole	A070909	Oversensing	10/03/2023	10/03/2023
									A0712	Pacing Problem	10/03/2023	10/03/2023
									A072201	High Impedance	10/03/2023	10/03/2023
									A090801	Signal Artifact/Noise	10/03/2023	10/03/2023
							E060106	Heart Block	A070909	Oversensing	10/03/2023	10/03/2023
									A0712	Pacing Problem	10/03/2023	10/03/2023
									A072201	High Impedance	10/03/2023	10/03/2023
									A090801	Signal Artifact/Noise	10/03/2023	10/03/2023
							F08	Hospitalization or Prolonged Hospitalization	A070909	Oversensing	10/03/2023	10/03/2023
									A0712	Pacing Problem	10/03/2023	10/03/2023
									A072201	High Impedance	10/03/2023	10/03/2023
									A090801	Signal Artifact/Noise	10/03/2023	10/03/2023
							F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A070909	Oversensing	10/03/2023	10/03/2023
									A0712	Pacing Problem	10/03/2023	10/03/2023
									A072201	High Impedance	10/03/2023	10/03/2023
									A090801	Signal Artifact/Noise	10/03/2023	10/03/2023
							F1203	Life Threatening Illness or Injury	A070909	Oversensing	10/03/2023	10/03/2023
									A0712	Pacing Problem	10/03/2023	10/03/2023
									A072201	High Impedance	10/03/2023	10/03/2023
									A090801	Signal Artifact/Noise	10/03/2023	10/03/2023
							F1905	Device Revision or Replacement	A070909	Oversensing	10/03/2023	10/03/2023
									A0712	Pacing Problem	10/03/2023	10/03/2023
									A072201	High Impedance	10/03/2023	10/03/2023
									A090801	Signal Artifact/Noise	10/03/2023	10/03/2023
					CT2406	Implantable	E011903	Syncope/Fainting	A070909	Oversensing	10/03/2023	10/03/2023
									A0712	Pacing Problem	10/03/2023	10/03/2023
									A072201	High Impedance	10/03/2023	10/03/2023
									A090801	Signal Artifact/Noise	10/03/2023	10/03/2023
							E060101	Asystole	A070909	Oversensing	10/03/2023	10/03/2023
									A0712	Pacing Problem	10/03/2023	10/03/2023
									A072201	High Impedance	10/03/2023	10/03/2023
									A090801	Signal Artifact/Noise	10/03/2023	10/03/2023
							E060106	Heart Block	A070909	Oversensing	10/03/2023	10/03/2023
									A0712	Pacing Problem	10/03/2023	10/03/2023
									A072201	High Impedance	10/03/2023	10/03/2023
									A090801	Signal Artifact/Noise	10/03/2023	10/03/2023
							F08	Hospitalization or Prolonged Hospitalization	A070909	Oversensing	10/03/2023	10/03/2023
									A0712	Pacing Problem	10/03/2023	10/03/2023
									A072201	High Impedance	10/03/2023	10/03/2023
									A090801	Signal Artifact/Noise	10/03/2023	10/03/2023
							F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A070909	Oversensing	10/03/2023	10/03/2023
									A0712	Pacing Problem	10/03/2023	10/03/2023
									A072201	High Impedance	10/03/2023	10/03/2023
									A090801	Signal Artifact/Noise	10/03/2023	10/03/2023
							F1203	Life Threatening Illness or Injury	A070909	Oversensing	10/03/2023	10/03/2023
									A0712	Pacing Problem	10/03/2023	10/03/2023
									A072201	High Impedance	10/03/2023	10/03/2023
									A090801	Signal Artifact/Noise	10/03/2023	10/03/2023
							F1905	Device Revision or Replacement	A070909	Oversensing	10/03/2023	10/03/2023
									A0712	Pacing Problem	10/03/2023	10/03/2023
									A072201	High Impedance	10/03/2023	10/03/2023
									A090801	Signal Artifact/Noise	10/03/2023	10/03/2023
					CT2407	Active implantable	E011903	Syncope/Fainting	A070909	Oversensing	10/03/2023	10/03/2023
									A0712	Pacing Problem	10/03/2023	10/03/2023
									A072201	High Impedance	10/03/2023	10/03/2023
									A090801	Signal Artifact/Noise	10/03/2023	10/03/2023
							E060101	Asystole	A070909	Oversensing	10/03/2023	10/03/2023
									A0712	Pacing Problem	10/03/2023	10/03/2023
									A072201	High Impedance	10/03/2023	10/03/2023
									A090801	Signal Artifact/Noise	10/03/2023	10/03/2023
							E060106	Heart Block	A070909	Oversensing	10/03/2023	10/03/2023
									A0712	Pacing Problem	10/03/2023	10/03/2023
									A072201	High Impedance	10/03/2023	10/03/2023
									A090801	Signal Artifact/Noise	10/03/2023	10/03/2023
							F08	Hospitalization or Prolonged Hospitalization	A070909	Oversensing	10/03/2023	10/03/2023
									A0712	Pacing Problem	10/03/2023	10/03/2023
									A072201	High Impedance	10/03/2023	10/03/2023

								A090801	Signal Artifact/Noise	10/03/2023	10/03/2023
							F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A070909	Oversensing	10/03/2023	10/03/2023
									A0712	Pacing Problem	10/03/2023	10/03/2023
									A072201	High Impedance	10/03/2023	10/03/2023
									A090801	Signal Artifact/Noise	10/03/2023	10/03/2023
							F1203	Life Threatening Illness or Injury	A070909	Oversensing	10/03/2023	10/03/2023
									A0712	Pacing Problem	10/03/2023	10/03/2023
									A072201	High Impedance	10/03/2023	10/03/2023
									A090801	Signal Artifact/Noise	10/03/2023	10/03/2023
							F1905	Device Revision or Replacement	A070909	Oversensing	10/03/2023	10/03/2023
									A0712	Pacing Problem	10/03/2023	10/03/2023
									A072201	High Impedance	10/03/2023	10/03/2023
									A090801	Signal Artifact/Noise	10/03/2023	10/03/2023
					CT257	Implantable cardiac pacemakers	E011903	Syncope/Fainting	A070909	Oversensing	10/03/2023	10/03/2023
									A0712	Pacing Problem	10/03/2023	10/03/2023
									A072201	High Impedance	10/03/2023	10/03/2023
									A090801	Signal Artifact/Noise	10/03/2023	10/03/2023
							E060101	Asystole	A070909	Oversensing	10/03/2023	10/03/2023
									A0712	Pacing Problem	10/03/2023	10/03/2023
									A072201	High Impedance	10/03/2023	10/03/2023
									A090801	Signal Artifact/Noise	10/03/2023	10/03/2023
							E060106	Heart Block	A070909	Oversensing	10/03/2023	10/03/2023
									A0712	Pacing Problem	10/03/2023	10/03/2023
									A072201	High Impedance	10/03/2023	10/03/2023
									A090801	Signal Artifact/Noise	10/03/2023	10/03/2023
							F08	Hospitalization or Prolonged Hospitalization	A070909	Oversensing	10/03/2023	10/03/2023
									A0712	Pacing Problem	10/03/2023	10/03/2023
									A072201	High Impedance	10/03/2023	10/03/2023
									A090801	Signal Artifact/Noise	10/03/2023	10/03/2023
							F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A070909	Oversensing	10/03/2023	10/03/2023
									A0712	Pacing Problem	10/03/2023	10/03/2023
									A072201	High Impedance	10/03/2023	10/03/2023
									A090801	Signal Artifact/Noise	10/03/2023	10/03/2023
							F1203	Life Threatening Illness or Injury	A070909	Oversensing	10/03/2023	10/03/2023
									A0712	Pacing Problem	10/03/2023	10/03/2023
									A072201	High Impedance	10/03/2023	10/03/2023
									A090801	Signal Artifact/Noise	10/03/2023	10/03/2023
							F1905	Device Revision or Replacement	A070909	Oversensing	10/03/2023	10/03/2023
									A0712	Pacing Problem	10/03/2023	10/03/2023
									A072201	High Impedance	10/03/2023	10/03/2023
									A090801	Signal Artifact/Noise	10/03/2023	10/03/2023
					CT2727	Active implant devices/ancillaries	E011903	Syncope/Fainting	A070909	Oversensing	10/03/2023	10/03/2023
									A0712	Pacing Problem	10/03/2023	10/03/2023
									A072201	High Impedance	10/03/2023	10/03/2023
									A090801	Signal Artifact/Noise	10/03/2023	10/03/2023
							E060101	Asystole	A070909	Oversensing	10/03/2023	10/03/2023
									A0712	Pacing Problem	10/03/2023	10/03/2023
									A072201	High Impedance	10/03/2023	10/03/2023
									A090801	Signal Artifact/Noise	10/03/2023	10/03/2023
							E060106	Heart Block	A070909	Oversensing	10/03/2023	10/03/2023
									A0712	Pacing Problem	10/03/2023	10/03/2023
									A072201	High Impedance	10/03/2023	10/03/2023
									A090801	Signal Artifact/Noise	10/03/2023	10/03/2023
							F08	Hospitalization or Prolonged Hospitalization	A070909	Oversensing	10/03/2023	10/03/2023
									A0712	Pacing Problem	10/03/2023	10/03/2023
									A072201	High Impedance	10/03/2023	10/03/2023
									A090801	Signal Artifact/Noise	10/03/2023	10/03/2023
							F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A070909	Oversensing	10/03/2023	10/03/2023
									A0712	Pacing Problem	10/03/2023	10/03/2023
									A072201	High Impedance	10/03/2023	10/03/2023
									A090801	Signal Artifact/Noise	10/03/2023	10/03/2023
							F1203	Life Threatening Illness or Injury	A070909	Oversensing	10/03/2023	10/03/2023
									A0712	Pacing Problem	10/03/2023	10/03/2023
									A072201	High Impedance	10/03/2023	10/03/2023
									A090801	Signal Artifact/Noise	10/03/2023	10/03/2023
							F1905	Device Revision or Replacement	A070909	Oversensing	10/03/2023	10/03/2023
									A0712	Pacing Problem	10/03/2023	10/03/2023
									A072201	High Impedance	10/03/2023	10/03/2023
									A090801	Signal Artifact/Noise	10/03/2023	10/03/2023
					CT2750	Implantable cardiac electrical stimulators and associated devices	E011903	Syncope/Fainting	A070909	Oversensing	10/03/2023	10/03/2023
									A0712	Pacing Problem	10/03/2023	10/03/2023
									A072201	High Impedance	10/03/2023	10/03/2023
									A090801	Signal Artifact/Noise	10/03/2023	10/03/2023
							E060101	Asystole	A070909	Oversensing	10/03/2023	10/03/2023
									A0712	Pacing Problem	10/03/2023	10/03/2023
									A072201	High Impedance	10/03/2023	10/03/2023
									A090801	Signal Artifact/Noise	10/03/2023	10/03/2023
							E060106	Heart Block	A070909	Oversensing	10/03/2023	10/03/2023
									A0712	Pacing Problem	10/03/2023	10/03/2023
									A072201	High Impedance	10/03/2023	10/03/2023
									A090801	Signal Artifact/Noise	10/03/2023	10/03/2023
							F08	Hospitalization or Prolonged Hospitalization	A070909	Oversensing	10/03/2023	10/03/2023
									A0712	Pacing Problem	10/03/2023	10/03/2023
									A072201	High Impedance	10/03/2023	10/03/2023
									A090801	Signal Artifact/Noise	10/03/2023	10/03/2023
							F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A070909	Oversensing	10/03/2023	10/03/2023
									A0712	Pacing Problem	10/03/2023	10/03/2023
									A072201	High Impedance	10/03/2023	10/03/2023

								A090801	Signal Artifact/Noise	10/03/2023	10/03/2023
						F1203	Life Threatening Illness or Injury	A070909	Oversensing	10/03/2023	10/03/2023
								A0712	Pacing Problem	10/03/2023	10/03/2023
								A072201	High Impedance	10/03/2023	10/03/2023
								A090801	Signal Artifact/Noise	10/03/2023	10/03/2023
						F1905	Device Revision or Replacement	A070909	Oversensing	10/03/2023	10/03/2023
								A0712	Pacing Problem	10/03/2023	10/03/2023
								A072201	High Impedance	10/03/2023	10/03/2023
								A090801	Signal Artifact/Noise	10/03/2023	10/03/2023
					CT2866	Implantable cardiac electrical stimulators	E011903	Syncope/Fainting	A070909	Oversensing	10/03/2023	10/03/2023
									A0712	Pacing Problem	10/03/2023	10/03/2023
									A072201	High Impedance	10/03/2023	10/03/2023
									A090801	Signal Artifact/Noise	10/03/2023	10/03/2023
							E060101	Asystole	A070909	Oversensing	10/03/2023	10/03/2023
									A0712	Pacing Problem	10/03/2023	10/03/2023
									A072201	High Impedance	10/03/2023	10/03/2023
									A090801	Signal Artifact/Noise	10/03/2023	10/03/2023
							E060106	Heart Block	A070909	Oversensing	10/03/2023	10/03/2023
									A0712	Pacing Problem	10/03/2023	10/03/2023
									A072201	High Impedance	10/03/2023	10/03/2023
									A090801	Signal Artifact/Noise	10/03/2023	10/03/2023
							F08	Hospitalization or Prolonged Hospitalization	A070909	Oversensing	10/03/2023	10/03/2023
									A0712	Pacing Problem	10/03/2023	10/03/2023
									A072201	High Impedance	10/03/2023	10/03/2023
									A090801	Signal Artifact/Noise	10/03/2023	10/03/2023
							F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A070909	Oversensing	10/03/2023	10/03/2023
									A0712	Pacing Problem	10/03/2023	10/03/2023
									A072201	High Impedance	10/03/2023	10/03/2023
									A090801	Signal Artifact/Noise	10/03/2023	10/03/2023
							F1203	Life Threatening Illness or Injury	A070909	Oversensing	10/03/2023	10/03/2023
									A0712	Pacing Problem	10/03/2023	10/03/2023
									A072201	High Impedance	10/03/2023	10/03/2023
									A090801	Signal Artifact/Noise	10/03/2023	10/03/2023
							F1905	Device Revision or Replacement	A070909	Oversensing	10/03/2023	10/03/2023
									A0712	Pacing Problem	10/03/2023	10/03/2023
									A072201	High Impedance	10/03/2023	10/03/2023
									A090801	Signal Artifact/Noise	10/03/2023	10/03/2023
					CT2985	Active	E011903	Syncope/Fainting	A070909	Oversensing	10/03/2023	10/03/2023
									A0712	Pacing Problem	10/03/2023	10/03/2023
									A072201	High Impedance	10/03/2023	10/03/2023
									A090801	Signal Artifact/Noise	10/03/2023	10/03/2023
							E060101	Asystole	A070909	Oversensing	10/03/2023	10/03/2023
									A0712	Pacing Problem	10/03/2023	10/03/2023
									A072201	High Impedance	10/03/2023	10/03/2023
									A090801	Signal Artifact/Noise	10/03/2023	10/03/2023
							E060106	Heart Block	A070909	Oversensing	10/03/2023	10/03/2023
									A0712	Pacing Problem	10/03/2023	10/03/2023
									A072201	High Impedance	10/03/2023	10/03/2023
									A090801	Signal Artifact/Noise	10/03/2023	10/03/2023
							F08	Hospitalization or Prolonged Hospitalization	A070909	Oversensing	10/03/2023	10/03/2023
									A0712	Pacing Problem	10/03/2023	10/03/2023
									A072201	High Impedance	10/03/2023	10/03/2023
									A090801	Signal Artifact/Noise	10/03/2023	10/03/2023
							F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A070909	Oversensing	10/03/2023	10/03/2023
									A0712	Pacing Problem	10/03/2023	10/03/2023
									A072201	High Impedance	10/03/2023	10/03/2023
									A090801	Signal Artifact/Noise	10/03/2023	10/03/2023
							F1203	Life Threatening Illness or Injury	A070909	Oversensing	10/03/2023	10/03/2023
									A0712	Pacing Problem	10/03/2023	10/03/2023
									A072201	High Impedance	10/03/2023	10/03/2023
									A090801	Signal Artifact/Noise	10/03/2023	10/03/2023
							F1905	Device Revision or Replacement	A070909	Oversensing	10/03/2023	10/03/2023
									A0712	Pacing Problem	10/03/2023	10/03/2023
									A072201	High Impedance	10/03/2023	10/03/2023
									A090801	Signal Artifact/Noise	10/03/2023	10/03/2023
					CT3001	Implantable metal	E011903	Syncope/Fainting	A070909	Oversensing	10/03/2023	10/03/2023
									A0712	Pacing Problem	10/03/2023	10/03/2023
									A072201	High Impedance	10/03/2023	10/03/2023
									A090801	Signal Artifact/Noise	10/03/2023	10/03/2023
							E060101	Asystole	A070909	Oversensing	10/03/2023	10/03/2023
									A0712	Pacing Problem	10/03/2023	10/03/2023
									A072201	High Impedance	10/03/2023	10/03/2023
									A090801	Signal Artifact/Noise	10/03/2023	10/03/2023
							E060106	Heart Block	A070909	Oversensing	10/03/2023	10/03/2023
									A0712	Pacing Problem	10/03/2023	10/03/2023
									A072201	High Impedance	10/03/2023	10/03/2023
									A090801	Signal Artifact/Noise	10/03/2023	10/03/2023
							F08	Hospitalization or Prolonged Hospitalization	A070909	Oversensing	10/03/2023	10/03/2023
									A0712	Pacing Problem	10/03/2023	10/03/2023
									A072201	High Impedance	10/03/2023	10/03/2023
									A090801	Signal Artifact/Noise	10/03/2023	10/03/2023
							F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A070909	Oversensing	10/03/2023	10/03/2023
									A0712	Pacing Problem	10/03/2023	10/03/2023
									A072201	High Impedance	10/03/2023	10/03/2023
									A090801	Signal Artifact/Noise	10/03/2023	10/03/2023
							F1203	Life Threatening Illness or Injury	A070909	Oversensing	10/03/2023	10/03/2023
									A0712	Pacing Problem	10/03/2023	10/03/2023
									A072201	High Impedance	10/03/2023	10/03/2023

								A090801	Signal Artifact/Noise	10/03/2023	10/03/2023
						F1905	Device Revision or Replacement	A070909	Oversensing	10/03/2023	10/03/2023
								A0712	Pacing Problem	10/03/2023	10/03/2023
								A072201	High Impedance	10/03/2023	10/03/2023
								A090801	Signal Artifact/Noise	10/03/2023	10/03/2023
					CT3002	Implantable synthetic polymer	E011903	Syncope/Fainting	A070909	Oversensing	10/03/2023	10/03/2023
									A0712	Pacing Problem	10/03/2023	10/03/2023
									A072201	High Impedance	10/03/2023	10/03/2023
									A090801	Signal Artifact/Noise	10/03/2023	10/03/2023
							E060101	Asystole	A070909	Oversensing	10/03/2023	10/03/2023
									A0712	Pacing Problem	10/03/2023	10/03/2023
									A072201	High Impedance	10/03/2023	10/03/2023
									A090801	Signal Artifact/Noise	10/03/2023	10/03/2023
							E060106	Heart Block	A070909	Oversensing	10/03/2023	10/03/2023
									A0712	Pacing Problem	10/03/2023	10/03/2023
									A072201	High Impedance	10/03/2023	10/03/2023
									A090801	Signal Artifact/Noise	10/03/2023	10/03/2023
							F08	Hospitalization or Prolonged Hospitalization	A070909	Oversensing	10/03/2023	10/03/2023
									A0712	Pacing Problem	10/03/2023	10/03/2023
									A072201	High Impedance	10/03/2023	10/03/2023
									A090801	Signal Artifact/Noise	10/03/2023	10/03/2023
							F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A070909	Oversensing	10/03/2023	10/03/2023
									A0712	Pacing Problem	10/03/2023	10/03/2023
									A072201	High Impedance	10/03/2023	10/03/2023
									A090801	Signal Artifact/Noise	10/03/2023	10/03/2023
							F1203	Life Threatening Illness or Injury	A070909	Oversensing	10/03/2023	10/03/2023
									A0712	Pacing Problem	10/03/2023	10/03/2023
									A072201	High Impedance	10/03/2023	10/03/2023
									A090801	Signal Artifact/Noise	10/03/2023	10/03/2023
							F1905	Device Revision or Replacement	A070909	Oversensing	10/03/2023	10/03/2023
									A0712	Pacing Problem	10/03/2023	10/03/2023
									A072201	High Impedance	10/03/2023	10/03/2023
									A090801	Signal Artifact/Noise	10/03/2023	10/03/2023
					CT3006	Implantable inorganic material	E011903	Syncope/Fainting	A070909	Oversensing	10/03/2023	10/03/2023
									A0712	Pacing Problem	10/03/2023	10/03/2023
									A072201	High Impedance	10/03/2023	10/03/2023
									A090801	Signal Artifact/Noise	10/03/2023	10/03/2023
							E060101	Asystole	A070909	Oversensing	10/03/2023	10/03/2023
									A0712	Pacing Problem	10/03/2023	10/03/2023
									A072201	High Impedance	10/03/2023	10/03/2023
									A090801	Signal Artifact/Noise	10/03/2023	10/03/2023
							E060106	Heart Block	A070909	Oversensing	10/03/2023	10/03/2023
									A0712	Pacing Problem	10/03/2023	10/03/2023
									A072201	High Impedance	10/03/2023	10/03/2023
									A090801	Signal Artifact/Noise	10/03/2023	10/03/2023
							F08	Hospitalization or Prolonged Hospitalization	A070909	Oversensing	10/03/2023	10/03/2023
									A0712	Pacing Problem	10/03/2023	10/03/2023
									A072201	High Impedance	10/03/2023	10/03/2023
									A090801	Signal Artifact/Noise	10/03/2023	10/03/2023
							F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A070909	Oversensing	10/03/2023	10/03/2023
									A0712	Pacing Problem	10/03/2023	10/03/2023
									A072201	High Impedance	10/03/2023	10/03/2023
									A090801	Signal Artifact/Noise	10/03/2023	10/03/2023
							F1203	Life Threatening Illness or Injury	A070909	Oversensing	10/03/2023	10/03/2023
									A0712	Pacing Problem	10/03/2023	10/03/2023
									A072201	High Impedance	10/03/2023	10/03/2023
									A090801	Signal Artifact/Noise	10/03/2023	10/03/2023
							F1905	Device Revision or Replacement	A070909	Oversensing	10/03/2023	10/03/2023
									A0712	Pacing Problem	10/03/2023	10/03/2023
									A072201	High Impedance	10/03/2023	10/03/2023
									A090801	Signal Artifact/Noise	10/03/2023	10/03/2023
					CT314	Electrically powered	E011903	Syncope/Fainting	A070909	Oversensing	10/03/2023	10/03/2023
									A0712	Pacing Problem	10/03/2023	10/03/2023
									A072201	High Impedance	10/03/2023	10/03/2023
									A090801	Signal Artifact/Noise	10/03/2023	10/03/2023
							E060101	Asystole	A070909	Oversensing	10/03/2023	10/03/2023
									A0712	Pacing Problem	10/03/2023	10/03/2023
									A072201	High Impedance	10/03/2023	10/03/2023
									A090801	Signal Artifact/Noise	10/03/2023	10/03/2023
							E060106	Heart Block	A070909	Oversensing	10/03/2023	10/03/2023
									A0712	Pacing Problem	10/03/2023	10/03/2023
									A072201	High Impedance	10/03/2023	10/03/2023
									A090801	Signal Artifact/Noise	10/03/2023	10/03/2023
							F08	Hospitalization or Prolonged Hospitalization	A070909	Oversensing	10/03/2023	10/03/2023
									A0712	Pacing Problem	10/03/2023	10/03/2023
									A072201	High Impedance	10/03/2023	10/03/2023
									A090801	Signal Artifact/Noise	10/03/2023	10/03/2023
							F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A070909	Oversensing	10/03/2023	10/03/2023
									A0712	Pacing Problem	10/03/2023	10/03/2023
									A072201	High Impedance	10/03/2023	10/03/2023
									A090801	Signal Artifact/Noise	10/03/2023	10/03/2023
							F1203	Life Threatening Illness or Injury	A070909	Oversensing	10/03/2023	10/03/2023
									A0712	Pacing Problem	10/03/2023	10/03/2023
									A072201	High Impedance	10/03/2023	10/03/2023
									A090801	Signal Artifact/Noise	10/03/2023	10/03/2023
							F1905	Device Revision or Replacement	A070909	Oversensing	10/03/2023	10/03/2023
									A0712	Pacing Problem	10/03/2023	10/03/2023
									A072201	High Impedance	10/03/2023	10/03/2023

								A090801	Signal Artifact/Noise	10/03/2023	10/03/2023
					CT321	Long-term surgical invasive	E011903	Syncope/Fainting	A070909	Oversensing	10/03/2023	10/03/2023
									A0712	Pacing Problem	10/03/2023	10/03/2023
									A072201	High Impedance	10/03/2023	10/03/2023
									A090801	Signal Artifact/Noise	10/03/2023	10/03/2023
							E060101	Asystole	A070909	Oversensing	10/03/2023	10/03/2023
									A0712	Pacing Problem	10/03/2023	10/03/2023
									A072201	High Impedance	10/03/2023	10/03/2023
									A090801	Signal Artifact/Noise	10/03/2023	10/03/2023
							E060106	Heart Block	A070909	Oversensing	10/03/2023	10/03/2023
									A0712	Pacing Problem	10/03/2023	10/03/2023
									A072201	High Impedance	10/03/2023	10/03/2023
									A090801	Signal Artifact/Noise	10/03/2023	10/03/2023
							F08	Hospitalization or Prolonged Hospitalization	A070909	Oversensing	10/03/2023	10/03/2023
									A0712	Pacing Problem	10/03/2023	10/03/2023
									A072201	High Impedance	10/03/2023	10/03/2023
									A090801	Signal Artifact/Noise	10/03/2023	10/03/2023
							F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A070909	Oversensing	10/03/2023	10/03/2023
									A0712	Pacing Problem	10/03/2023	10/03/2023
									A072201	High Impedance	10/03/2023	10/03/2023
									A090801	Signal Artifact/Noise	10/03/2023	10/03/2023
							F1203	Life Threatening Illness or Injury	A070909	Oversensing	10/03/2023	10/03/2023
									A0712	Pacing Problem	10/03/2023	10/03/2023
									A072201	High Impedance	10/03/2023	10/03/2023
									A090801	Signal Artifact/Noise	10/03/2023	10/03/2023
							F1905	Device Revision or Replacement	A070909	Oversensing	10/03/2023	10/03/2023
									A0712	Pacing Problem	10/03/2023	10/03/2023
									A072201	High Impedance	10/03/2023	10/03/2023
									A090801	Signal Artifact/Noise	10/03/2023	10/03/2023
					CT449	Generators and associated devices	E011903	Syncope/Fainting	A070909	Oversensing	10/03/2023	10/03/2023
									A0712	Pacing Problem	10/03/2023	10/03/2023
									A072201	High Impedance	10/03/2023	10/03/2023
									A090801	Signal Artifact/Noise	10/03/2023	10/03/2023
							E060101	Asystole	A070909	Oversensing	10/03/2023	10/03/2023
									A0712	Pacing Problem	10/03/2023	10/03/2023
									A072201	High Impedance	10/03/2023	10/03/2023
									A090801	Signal Artifact/Noise	10/03/2023	10/03/2023
							E060106	Heart Block	A070909	Oversensing	10/03/2023	10/03/2023
									A0712	Pacing Problem	10/03/2023	10/03/2023
									A072201	High Impedance	10/03/2023	10/03/2023
									A090801	Signal Artifact/Noise	10/03/2023	10/03/2023
							F08	Hospitalization or Prolonged Hospitalization	A070909	Oversensing	10/03/2023	10/03/2023
									A0712	Pacing Problem	10/03/2023	10/03/2023
									A072201	High Impedance	10/03/2023	10/03/2023
									A090801	Signal Artifact/Noise	10/03/2023	10/03/2023
							F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A070909	Oversensing	10/03/2023	10/03/2023
									A0712	Pacing Problem	10/03/2023	10/03/2023
									A072201	High Impedance	10/03/2023	10/03/2023
									A090801	Signal Artifact/Noise	10/03/2023	10/03/2023
							F1203	Life Threatening Illness or Injury	A070909	Oversensing	10/03/2023	10/03/2023
									A0712	Pacing Problem	10/03/2023	10/03/2023
									A072201	High Impedance	10/03/2023	10/03/2023
									A090801	Signal Artifact/Noise	10/03/2023	10/03/2023
							F1905	Device Revision or Replacement	A070909	Oversensing	10/03/2023	10/03/2023
									A0712	Pacing Problem	10/03/2023	10/03/2023
									A072201	High Impedance	10/03/2023	10/03/2023
									A090801	Signal Artifact/Noise	10/03/2023	10/03/2023
					CT981	Single use	E011903	Syncope/Fainting	A070909	Oversensing	10/03/2023	10/03/2023
									A0712	Pacing Problem	10/03/2023	10/03/2023
									A072201	High Impedance	10/03/2023	10/03/2023
									A090801	Signal Artifact/Noise	10/03/2023	10/03/2023
							E060101	Asystole	A070909	Oversensing	10/03/2023	10/03/2023
									A0712	Pacing Problem	10/03/2023	10/03/2023
									A072201	High Impedance	10/03/2023	10/03/2023
									A090801	Signal Artifact/Noise	10/03/2023	10/03/2023
							E060106	Heart Block	A070909	Oversensing	10/03/2023	10/03/2023
									A0712	Pacing Problem	10/03/2023	10/03/2023
									A072201	High Impedance	10/03/2023	10/03/2023
									A090801	Signal Artifact/Noise	10/03/2023	10/03/2023
							F08	Hospitalization or Prolonged Hospitalization	A070909	Oversensing	10/03/2023	10/03/2023
									A0712	Pacing Problem	10/03/2023	10/03/2023
									A072201	High Impedance	10/03/2023	10/03/2023
									A090801	Signal Artifact/Noise	10/03/2023	10/03/2023
							F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A070909	Oversensing	10/03/2023	10/03/2023
									A0712	Pacing Problem	10/03/2023	10/03/2023
									A072201	High Impedance	10/03/2023	10/03/2023
									A090801	Signal Artifact/Noise	10/03/2023	10/03/2023
							F1203	Life Threatening Illness or Injury	A070909	Oversensing	10/03/2023	10/03/2023
									A0712	Pacing Problem	10/03/2023	10/03/2023
									A072201	High Impedance	10/03/2023	10/03/2023
									A090801	Signal Artifact/Noise	10/03/2023	10/03/2023
							F1905	Device Revision or Replacement	A070909	Oversensing	10/03/2023	10/03/2023
									A0712	Pacing Problem	10/03/2023	10/03/2023
									A072201	High Impedance	10/03/2023	10/03/2023
									A090801	Signal Artifact/Noise	10/03/2023	10/03/2023
					CT983	Surgical invasive	E011903	Syncope/Fainting	A070909	Oversensing	10/03/2023	10/03/2023
									A0712	Pacing Problem	10/03/2023	10/03/2023
									A072201	High Impedance	10/03/2023	10/03/2023

								A090801	Signal Artifact/Noise	10/03/2023	10/03/2023
						E060101	Asystole	A070909	Oversensing	10/03/2023	10/03/2023
								A0712	Pacing Problem	10/03/2023	10/03/2023
								A072201	High Impedance	10/03/2023	10/03/2023
								A090801	Signal Artifact/Noise	10/03/2023	10/03/2023
						E060106	Heart Block	A070909	Oversensing	10/03/2023	10/03/2023
								A0712	Pacing Problem	10/03/2023	10/03/2023
								A072201	High Impedance	10/03/2023	10/03/2023
								A090801	Signal Artifact/Noise	10/03/2023	10/03/2023
						F08	Hospitalization or Prolonged Hospitalization	A070909	Oversensing	10/03/2023	10/03/2023
								A0712	Pacing Problem	10/03/2023	10/03/2023
								A072201	High Impedance	10/03/2023	10/03/2023
								A090801	Signal Artifact/Noise	10/03/2023	10/03/2023
						F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A070909	Oversensing	10/03/2023	10/03/2023
								A0712	Pacing Problem	10/03/2023	10/03/2023
								A072201	High Impedance	10/03/2023	10/03/2023
								A090801	Signal Artifact/Noise	10/03/2023	10/03/2023
						F1203	Life Threatening Illness or Injury	A070909	Oversensing	10/03/2023	10/03/2023
								A0712	Pacing Problem	10/03/2023	10/03/2023
								A072201	High Impedance	10/03/2023	10/03/2023
								A090801	Signal Artifact/Noise	10/03/2023	10/03/2023
						F1905	Device Revision or Replacement	A070909	Oversensing	10/03/2023	10/03/2023
								A0712	Pacing Problem	10/03/2023	10/03/2023
								A072201	High Impedance	10/03/2023	10/03/2023
								A090801	Signal Artifact/Noise	10/03/2023	10/03/2023
12832	2023	All other reportable incidents	47265	Dual-chamber implantable pacemaker, rate-responsive	CT1685	Cardiac pacemakers	E011903	Syncope/Fainting	A070909	Oversensing	10/03/2023	10/03/2023
									A0712	Pacing Problem	10/03/2023	10/03/2023
									A072201	High Impedance	10/03/2023	10/03/2023
									A090801	Signal Artifact/Noise	10/03/2023	10/03/2023
							E060101	Asystole	A070909	Oversensing	10/03/2023	10/03/2023
									A0712	Pacing Problem	10/03/2023	10/03/2023
									A072201	High Impedance	10/03/2023	10/03/2023
									A090801	Signal Artifact/Noise	10/03/2023	10/03/2023
							E060106	Heart Block	A070909	Oversensing	10/03/2023	10/03/2023
									A0712	Pacing Problem	10/03/2023	10/03/2023
									A072201	High Impedance	10/03/2023	10/03/2023
									A090801	Signal Artifact/Noise	10/03/2023	10/03/2023
							F08	Hospitalization or Prolonged Hospitalization	A070909	Oversensing	10/03/2023	10/03/2023
									A0712	Pacing Problem	10/03/2023	10/03/2023
									A072201	High Impedance	10/03/2023	10/03/2023
									A090801	Signal Artifact/Noise	10/03/2023	10/03/2023
							F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A070909	Oversensing	10/03/2023	10/03/2023
									A0712	Pacing Problem	10/03/2023	10/03/2023
									A072201	High Impedance	10/03/2023	10/03/2023
									A090801	Signal Artifact/Noise	10/03/2023	10/03/2023
							F1203	Life Threatening Illness or Injury	A070909	Oversensing	10/03/2023	10/03/2023
									A0712	Pacing Problem	10/03/2023	10/03/2023
									A072201	High Impedance	10/03/2023	10/03/2023
									A090801	Signal Artifact/Noise	10/03/2023	10/03/2023
							F1905	Device Revision or Replacement	A070909	Oversensing	10/03/2023	10/03/2023
									A0712	Pacing Problem	10/03/2023	10/03/2023
									A072201	High Impedance	10/03/2023	10/03/2023
									A090801	Signal Artifact/Noise	10/03/2023	10/03/2023
					CT257	Implantable cardiac pacemakers	E011903	Syncope/Fainting	A070909	Oversensing	10/03/2023	10/03/2023
									A0712	Pacing Problem	10/03/2023	10/03/2023
									A072201	High Impedance	10/03/2023	10/03/2023
									A090801	Signal Artifact/Noise	10/03/2023	10/03/2023
							E060101	Asystole	A070909	Oversensing	10/03/2023	10/03/2023
									A0712	Pacing Problem	10/03/2023	10/03/2023
									A072201	High Impedance	10/03/2023	10/03/2023
									A090801	Signal Artifact/Noise	10/03/2023	10/03/2023
							E060106	Heart Block	A070909	Oversensing	10/03/2023	10/03/2023
									A0712	Pacing Problem	10/03/2023	10/03/2023
									A072201	High Impedance	10/03/2023	10/03/2023
									A090801	Signal Artifact/Noise	10/03/2023	10/03/2023
							F08	Hospitalization or Prolonged Hospitalization	A070909	Oversensing	10/03/2023	10/03/2023
									A0712	Pacing Problem	10/03/2023	10/03/2023
									A072201	High Impedance	10/03/2023	10/03/2023
									A090801	Signal Artifact/Noise	10/03/2023	10/03/2023
							F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A070909	Oversensing	10/03/2023	10/03/2023
									A0712	Pacing Problem	10/03/2023	10/03/2023
									A072201	High Impedance	10/03/2023	10/03/2023
									A090801	Signal Artifact/Noise	10/03/2023	10/03/2023
							F1203	Life Threatening Illness or Injury	A070909	Oversensing	10/03/2023	10/03/2023
									A0712	Pacing Problem	10/03/2023	10/03/2023
									A072201	High Impedance	10/03/2023	10/03/2023
									A090801	Signal Artifact/Noise	10/03/2023	10/03/2023
							F1905	Device Revision or Replacement	A070909	Oversensing	10/03/2023	10/03/2023
									A0712	Pacing Problem	10/03/2023	10/03/2023
									A072201	High Impedance	10/03/2023	10/03/2023
									A090801	Signal Artifact/Noise	10/03/2023	10/03/2023
					CT2866	Implantable cardiac electrical stimulators	E011903	Syncope/Fainting	A070909	Oversensing	10/03/2023	10/03/2023
									A0712	Pacing Problem	10/03/2023	10/03/2023
									A072201	High Impedance	10/03/2023	10/03/2023
									A090801	Signal Artifact/Noise	10/03/2023	10/03/2023
							E060101	Asystole	A070909	Oversensing	10/03/2023	10/03/2023
									A0712	Pacing Problem	10/03/2023	10/03/2023
									A072201	High Impedance	10/03/2023	10/03/2023

									A090801	Signal Artifact/Noise	10/03/2023	10/03/2023
							E060106	Heart Block	A070909	Oversensing	10/03/2023	10/03/2023
									A0712	Pacing Problem	10/03/2023	10/03/2023
									A072201	High Impedance	10/03/2023	10/03/2023
									A090801	Signal Artifact/Noise	10/03/2023	10/03/2023
							F08	Hospitalization or Prolonged Hospitalization	A070909	Oversensing	10/03/2023	10/03/2023
									A0712	Pacing Problem	10/03/2023	10/03/2023
									A072201	High Impedance	10/03/2023	10/03/2023
									A090801	Signal Artifact/Noise	10/03/2023	10/03/2023
							F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A070909	Oversensing	10/03/2023	10/03/2023
									A0712	Pacing Problem	10/03/2023	10/03/2023
									A072201	High Impedance	10/03/2023	10/03/2023
									A090801	Signal Artifact/Noise	10/03/2023	10/03/2023
							F1203	Life Threatening Illness or Injury	A070909	Oversensing	10/03/2023	10/03/2023
									A0712	Pacing Problem	10/03/2023	10/03/2023
									A072201	High Impedance	10/03/2023	10/03/2023
									A090801	Signal Artifact/Noise	10/03/2023	10/03/2023
							F1905	Device Revision or Replacement	A070909	Oversensing	10/03/2023	10/03/2023
									A0712	Pacing Problem	10/03/2023	10/03/2023
									A072201	High Impedance	10/03/2023	10/03/2023
									A090801	Signal Artifact/Noise	10/03/2023	10/03/2023
12833	2023		47265	Dual-chamber implantable pacemaker, rate-responsive	CT141	Electrodes/Leads and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	24/02/2023	24/02/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	24/02/2023	24/02/2023
					CT145	Cardiovascular devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	24/02/2023	24/02/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	24/02/2023	24/02/2023
					CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	24/02/2023	24/02/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	24/02/2023	24/02/2023
					CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	24/02/2023	24/02/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	24/02/2023	24/02/2023
					CT233	Surgical	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	24/02/2023	24/02/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	24/02/2023	24/02/2023
					CT2406	Implantable	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	24/02/2023	24/02/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	24/02/2023	24/02/2023
					CT2407	Active implantable	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	24/02/2023	24/02/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	24/02/2023	24/02/2023
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	24/02/2023	24/02/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	24/02/2023	24/02/2023
					CT2727	Active implant devices/ancillaries	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	24/02/2023	24/02/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	24/02/2023	24/02/2023
					CT2750	Implantable cardiac electrical stimulators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	24/02/2023	24/02/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	24/02/2023	24/02/2023
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	24/02/2023	24/02/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	24/02/2023	24/02/2023
					CT2985	Active	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	24/02/2023	24/02/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	24/02/2023	24/02/2023
					CT3001	Implantable metal	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	24/02/2023	24/02/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	24/02/2023	24/02/2023
					CT3002	Implantable synthetic polymer	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	24/02/2023	24/02/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	24/02/2023	24/02/2023
					CT3006	Implantable inorganic material	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	24/02/2023	24/02/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	24/02/2023	24/02/2023
					CT314	Electrically powered	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	24/02/2023	24/02/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	24/02/2023	24/02/2023
					CT321	Long-term surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	24/02/2023	24/02/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	24/02/2023	24/02/2023
					CT449	Generators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	24/02/2023	24/02/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	24/02/2023	24/02/2023
					CT981	Single use	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	24/02/2023	24/02/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	24/02/2023	24/02/2023
					CT983	Surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	24/02/2023	24/02/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	24/02/2023	24/02/2023
12834	2023		47265	Dual-chamber implantable pacemaker, rate-responsive	CT141	Electrodes/Leads and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	13/12/2022	13/12/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	13/12/2022	13/12/2022
					CT145	Cardiovascular devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	13/12/2022	13/12/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	13/12/2022	13/12/2022
					CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	13/12/2022	13/12/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	13/12/2022	13/12/2022
					CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	13/12/2022	13/12/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	13/12/2022	13/12/2022
					CT233	Surgical	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	13/12/2022	13/12/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	13/12/2022	13/12/2022
					CT2406	Implantable	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	13/12/2022	13/12/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	13/12/2022	13/12/2022
					CT2407	Active implantable	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	13/12/2022	13/12/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	13/12/2022	13/12/2022
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	13/12/2022	13/12/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	13/12/2022	13/12/2022
					CT2727	Active implant devices/ancillaries	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	13/12/2022	13/12/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	13/12/2022	13/12/2022
					CT2750	Implantable cardiac electrical stimulators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	13/12/2022	13/12/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	13/12/2022	13/12/2022
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	13/12/2022	13/12/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	13/12/2022	13/12/2022
					CT2985	Active	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	13/12/2022	13/12/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	13/12/2022	13/12/2022
					CT3001	Implantable metal	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	13/12/2022	13/12/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	13/12/2022	13/12/2022
					CT3002	Implantable synthetic polymer	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	13/12/2022	13/12/2022

							F1901	Additional Surgery	A24	No Apparent Device Problem	13/12/2022	13/12/2022
					CT3006	Implantable inorganic material	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	13/12/2022	13/12/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	13/12/2022	13/12/2022
					CT314	Electrically powered	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	13/12/2022	13/12/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	13/12/2022	13/12/2022
					CT321	Long-term surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	13/12/2022	13/12/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	13/12/2022	13/12/2022
					CT449	Generators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	13/12/2022	13/12/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	13/12/2022	13/12/2022
					CT981	Single use	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	13/12/2022	13/12/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	13/12/2022	13/12/2022
					CT983	Surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	13/12/2022	13/12/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	13/12/2022	13/12/2022
12835	2023	All other reportable incidents	47265	Dual-chamber implantable pacemaker, rate-responsive	CT141	Electrodes/Leads and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A0908	Incorrect, Inadequate or Imprecise Result or Readings	03/03/2023	03/03/2023
							F1901	Additional Surgery	A0908	Incorrect, Inadequate or Imprecise Result or Readings	03/03/2023	03/03/2023
					CT145	Cardiovascular devices	E2403	No Clinical Signs, Symptoms or Conditions	A0908	Incorrect, Inadequate or Imprecise Result or Readings	03/03/2023	03/03/2023
							F1901	Additional Surgery	A0908	Incorrect, Inadequate or Imprecise Result or Readings	03/03/2023	03/03/2023
					CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A0908	Incorrect, Inadequate or Imprecise Result or Readings	03/03/2023	03/03/2023
							F1901	Additional Surgery	A0908	Incorrect, Inadequate or Imprecise Result or Readings	03/03/2023	03/03/2023
					CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A0908	Incorrect, Inadequate or Imprecise Result or Readings	03/03/2023	03/03/2023
							F1901	Additional Surgery	A0908	Incorrect, Inadequate or Imprecise Result or Readings	03/03/2023	03/03/2023
					CT233	Surgical	E2403	No Clinical Signs, Symptoms or Conditions	A0908	Incorrect, Inadequate or Imprecise Result or Readings	03/03/2023	03/03/2023
							F1901	Additional Surgery	A0908	Incorrect, Inadequate or Imprecise Result or Readings	03/03/2023	03/03/2023
					CT2406	Implantable	E2403	No Clinical Signs, Symptoms or Conditions	A0908	Incorrect, Inadequate or Imprecise Result or Readings	03/03/2023	03/03/2023
							F1901	Additional Surgery	A0908	Incorrect, Inadequate or Imprecise Result or Readings	03/03/2023	03/03/2023
					CT2407	Active implantable	E2403	No Clinical Signs, Symptoms or Conditions	A0908	Incorrect, Inadequate or Imprecise Result or Readings	03/03/2023	03/03/2023
							F1901	Additional Surgery	A0908	Incorrect, Inadequate or Imprecise Result or Readings	03/03/2023	03/03/2023
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A0908	Incorrect, Inadequate or Imprecise Result or Readings	03/03/2023	03/03/2023
							F1901	Additional Surgery	A0908	Incorrect, Inadequate or Imprecise Result or Readings	03/03/2023	03/03/2023
					CT2727	Active implant devices/ancillaries	E2403	No Clinical Signs, Symptoms or Conditions	A0908	Incorrect, Inadequate or Imprecise Result or Readings	03/03/2023	03/03/2023
							F1901	Additional Surgery	A0908	Incorrect, Inadequate or Imprecise Result or Readings	03/03/2023	03/03/2023
					CT2750	Implantable cardiac electrical stimulators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A0908	Incorrect, Inadequate or Imprecise Result or Readings	03/03/2023	03/03/2023
							F1901	Additional Surgery	A0908	Incorrect, Inadequate or Imprecise Result or Readings	03/03/2023	03/03/2023
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A0908	Incorrect, Inadequate or Imprecise Result or Readings	03/03/2023	03/03/2023
							F1901	Additional Surgery	A0908	Incorrect, Inadequate or Imprecise Result or Readings	03/03/2023	03/03/2023
					CT2985	Active	E2403	No Clinical Signs, Symptoms or Conditions	A0908	Incorrect, Inadequate or Imprecise Result or Readings	03/03/2023	03/03/2023
							F1901	Additional Surgery	A0908	Incorrect, Inadequate or Imprecise Result or Readings	03/03/2023	03/03/2023
					CT3001	Implantable metal	E2403	No Clinical Signs, Symptoms or Conditions	A0908	Incorrect, Inadequate or Imprecise Result or Readings	03/03/2023	03/03/2023
							F1901	Additional Surgery	A0908	Incorrect, Inadequate or Imprecise Result or Readings	03/03/2023	03/03/2023
					CT3002	Implantable synthetic polymer	E2403	No Clinical Signs, Symptoms or Conditions	A0908	Incorrect, Inadequate or Imprecise Result or Readings	03/03/2023	03/03/2023
							F1901	Additional Surgery	A0908	Incorrect, Inadequate or Imprecise Result or Readings	03/03/2023	03/03/2023
					CT3006	Implantable inorganic material	E2403	No Clinical Signs, Symptoms or Conditions	A0908	Incorrect, Inadequate or Imprecise Result or Readings	03/03/2023	03/03/2023
							F1901	Additional Surgery	A0908	Incorrect, Inadequate or Imprecise Result or Readings	03/03/2023	03/03/2023
					CT314	Electrically powered	E2403	No Clinical Signs, Symptoms or Conditions	A0908	Incorrect, Inadequate or Imprecise Result or Readings	03/03/2023	03/03/2023
							F1901	Additional Surgery	A0908	Incorrect, Inadequate or Imprecise Result or Readings	03/03/2023	03/03/2023
					CT321	Long-term surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A0908	Incorrect, Inadequate or Imprecise Result or Readings	03/03/2023	03/03/2023
							F1901	Additional Surgery	A0908	Incorrect, Inadequate or Imprecise Result or Readings	03/03/2023	03/03/2023
					CT449	Generators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A0908	Incorrect, Inadequate or Imprecise Result or Readings	03/03/2023	03/03/2023
							F1901	Additional Surgery	A0908	Incorrect, Inadequate or Imprecise Result or Readings	03/03/2023	03/03/2023
					CT981	Single use	E2403	No Clinical Signs, Symptoms or Conditions	A0908	Incorrect, Inadequate or Imprecise Result or Readings	03/03/2023	03/03/2023
							F1901	Additional Surgery	A0908	Incorrect, Inadequate or Imprecise Result or Readings	03/03/2023	03/03/2023
					CT983	Surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A0908	Incorrect, Inadequate or Imprecise Result or Readings	03/03/2023	03/03/2023
							F1901	Additional Surgery	A0908	Incorrect, Inadequate or Imprecise Result or Readings	03/03/2023	03/03/2023
12836	2023	All other reportable incidents	47263	Cardiac resynchronization therapy implantable pacemaker	CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A090805	Incorrect Measurement	26/04/2023	26/04/2023
							F1901	Additional Surgery	A090805	Incorrect Measurement	26/04/2023	26/04/2023
					CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A090805	Incorrect Measurement	26/04/2023	26/04/2023
							F1901	Additional Surgery	A090805	Incorrect Measurement	26/04/2023	26/04/2023

					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A090805	Incorrect Measurement	26/04/2023	26/04/2023
							F1901	Additional Surgery	A090805	Incorrect Measurement	26/04/2023	26/04/2023
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A090805	Incorrect Measurement	26/04/2023	26/04/2023
							F1901	Additional Surgery	A090805	Incorrect Measurement	26/04/2023	26/04/2023
12837	2023	All other reportable incidents	35852	Single-chamber implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A051201	Device Dislodged or Dislocated		
									A0722	Impedance Problem		
							F08	Hospitalization or Prolonged Hospitalization	A051201	Device Dislodged or Dislocated		
									A0722	Impedance Problem		
							F1905	Device Revision or Replacement	A051201	Device Dislodged or Dislocated		
									A0722	Impedance Problem		
12838	2023	All other reportable incidents	47270	Cardiac resynchronization therapy implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A0722	Impedance Problem		
							F26	No Health Consequences or Impact	A0722	Impedance Problem		
12839	2023	All other reportable incidents	37265	Dual-chamber implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E060109	Tachycardia	A0709	Device Sensing Problem		
									A071301	Failure to Deliver Shock/Stimulation		
							F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A0709	Device Sensing Problem		
									A071301	Failure to Deliver Shock/Stimulation		
							F1001	Absence of Treatment	A0709	Device Sensing Problem		
									A071301	Failure to Deliver Shock/Stimulation		
							F1203	Life Threatening Illness or Injury	A0709	Device Sensing Problem		
									A071301	Failure to Deliver Shock/Stimulation		
							F23	Unexpected Medical Intervention	A0709	Device Sensing Problem		
									A071301	Failure to Deliver Shock/Stimulation		
12840	2023	All other reportable incidents	47265	Dual-chamber implantable pacemaker, rate-responsive	CT141	Electrodes/Leads and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	18/01/2023	18/01/2023
							F1901	Additional Surgery	A070504	Premature Discharge of Battery	18/01/2023	18/01/2023
					CT145	Cardiovascular devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	18/01/2023	18/01/2023
							F1901	Additional Surgery	A070504	Premature Discharge of Battery	18/01/2023	18/01/2023
					CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	18/01/2023	18/01/2023
							F1901	Additional Surgery	A070504	Premature Discharge of Battery	18/01/2023	18/01/2023
					CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	18/01/2023	18/01/2023
							F1901	Additional Surgery	A070504	Premature Discharge of Battery	18/01/2023	18/01/2023
					CT233	Surgical	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	18/01/2023	18/01/2023
							F1901	Additional Surgery	A070504	Premature Discharge of Battery	18/01/2023	18/01/2023
					CT2406	Implantable	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	18/01/2023	18/01/2023
							F1901	Additional Surgery	A070504	Premature Discharge of Battery	18/01/2023	18/01/2023
					CT2407	Active implantable	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	18/01/2023	18/01/2023
							F1901	Additional Surgery	A070504	Premature Discharge of Battery	18/01/2023	18/01/2023
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	18/01/2023	18/01/2023
							F1901	Additional Surgery	A070504	Premature Discharge of Battery	18/01/2023	18/01/2023
					CT2727	Active implant devices/ancillaries	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	18/01/2023	18/01/2023
							F1901	Additional Surgery	A070504	Premature Discharge of Battery	18/01/2023	18/01/2023
					CT2750	Implantable cardiac electrical stimulators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	18/01/2023	18/01/2023
							F1901	Additional Surgery	A070504	Premature Discharge of Battery	18/01/2023	18/01/2023
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	18/01/2023	18/01/2023
							F1901	Additional Surgery	A070504	Premature Discharge of Battery	18/01/2023	18/01/2023
					CT2985	Active	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	18/01/2023	18/01/2023
							F1901	Additional Surgery	A070504	Premature Discharge of Battery	18/01/2023	18/01/2023
					CT3001	Implantable metal	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	18/01/2023	18/01/2023
							F1901	Additional Surgery	A070504	Premature Discharge of Battery	18/01/2023	18/01/2023
					CT3002	Implantable synthetic polymer	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	18/01/2023	18/01/2023
							F1901	Additional Surgery	A070504	Premature Discharge of Battery	18/01/2023	18/01/2023
					CT3006	Implantable inorganic material	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	18/01/2023	18/01/2023
							F1901	Additional Surgery	A070504	Premature Discharge of Battery	18/01/2023	18/01/2023
					CT314	Electrically powered	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	18/01/2023	18/01/2023
							F1901	Additional Surgery	A070504	Premature Discharge of Battery	18/01/2023	18/01/2023
					CT321	Long-term surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	18/01/2023	18/01/2023
							F1901	Additional Surgery	A070504	Premature Discharge of Battery	18/01/2023	18/01/2023
					CT449	Generators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	18/01/2023	18/01/2023
							F1901	Additional Surgery	A070504	Premature Discharge of Battery	18/01/2023	18/01/2023
					CT981	Single use	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	18/01/2023	18/01/2023
							F1901	Additional Surgery	A070504	Premature Discharge of Battery	18/01/2023	18/01/2023
					CT983	Surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	18/01/2023	18/01/2023
							F1901	Additional Surgery	A070504	Premature Discharge of Battery	18/01/2023	18/01/2023
12841	2023	All other reportable incidents	17882	Physiologic-monitoring defibrillation system	CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A0719	Unexpected Shutdown		
							F27	Problem Identified During Non-Clinical Procedure	A0719	Unexpected Shutdown		
12842	2023	All other reportable incidents	47265	Dual-chamber implantable pacemaker, rate-responsive	CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem		
									A090805	Incorrect Measurement		
							F26	No Health Consequences or Impact	A0705	Battery Problem		
									A090805	Incorrect Measurement		
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem		
									A090805	Incorrect Measurement		
							F26	No Health Consequences or Impact	A0705	Battery Problem		
									A090805	Incorrect Measurement		
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem		
									A090805	Incorrect Measurement		
							F26	No Health Consequences or Impact	A0705	Battery Problem		
									A090805	Incorrect Measurement		
12843	2023	All other reportable incidents	36241	Pacemaker/defibrillator lead	CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070102	High Capture Threshold		
									A070909	Oversensing		
									A090801	Signal Artifact/Noise		
									A1301	Failure to Read Input Signal		
							F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A070102	High Capture Threshold		
									A070909	Oversensing		
									A090801	Signal Artifact/Noise		
									A1301	Failure to Read Input Signal		
12844	2023	All other reportable incidents	47263	Cardiac resynchronization therapy implantable pacemaker	CT1684	Cardiac pacemakers and associated devices	E2103	Device Overstimulation of Tissue	A071209	Extracardiac Stimulation		
									A0908	Incorrect, Inadequate or Imprecise Result or Readings		
							F08	Hospitalization or Prolonged Hospitalization	A071209	Extracardiac Stimulation		

									A0908	Incorrect, Inadequate or Imprecise Result or Readings		
					CT1685	Cardiac pacemakers	E2103	Device Overstimulation of Tissue	A071209	Extracardiac Stimulation		
									A0908	Incorrect, Inadequate or Imprecise Result or Readings		
							F08	Hospitalization or Prolonged Hospitalization	A071209	Extracardiac Stimulation		
									A0908	Incorrect, Inadequate or Imprecise Result or Readings		
					CT257	Implantable cardiac pacemakers	E2103	Device Overstimulation of Tissue	A071209	Extracardiac Stimulation		
									A0908	Incorrect, Inadequate or Imprecise Result or Readings		
							F08	Hospitalization or Prolonged Hospitalization	A071209	Extracardiac Stimulation		
									A0908	Incorrect, Inadequate or Imprecise Result or Readings		
					CT2866	Implantable cardiac electrical stimulators	E2103	Device Overstimulation of Tissue	A071209	Extracardiac Stimulation		
									A0908	Incorrect, Inadequate or Imprecise Result or Readings		
							F08	Hospitalization or Prolonged Hospitalization	A071209	Extracardiac Stimulation		
									A0908	Incorrect, Inadequate or Imprecise Result or Readings		
12845	2023	All other reportable incidents	17882	Physiologic-monitoring defibrillation system	CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A0719	Unexpected Shutdown		
							F27	Problem Identified During Non-Clinical Procedure	A0719	Unexpected Shutdown		
12846	2023	All other reportable incidents	47270	Cardiac resynchronization therapy implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A070909	Oversensing		
							F26	No Health Consequences or Impact	A070909	Oversensing		
12847	2023	All other reportable incidents	47265	Dual-chamber implantable pacemaker, rate-responsive	CT141	Electrodes/Leads and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	17/02/2023	17/02/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	17/02/2023	17/02/2023
					CT145	Cardiovascular devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	17/02/2023	17/02/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	17/02/2023	17/02/2023
					CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	17/02/2023	17/02/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	17/02/2023	17/02/2023
					CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	17/02/2023	17/02/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	17/02/2023	17/02/2023
					CT233	Surgical	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	17/02/2023	17/02/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	17/02/2023	17/02/2023
					CT2406	Implantable	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	17/02/2023	17/02/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	17/02/2023	17/02/2023
					CT2407	Active implantable	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	17/02/2023	17/02/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	17/02/2023	17/02/2023
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	17/02/2023	17/02/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	17/02/2023	17/02/2023
					CT2727	Active implant devices/ancillaries	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	17/02/2023	17/02/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	17/02/2023	17/02/2023
					CT2750	Implantable cardiac electrical stimulators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	17/02/2023	17/02/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	17/02/2023	17/02/2023
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	17/02/2023	17/02/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	17/02/2023	17/02/2023
					CT2985	Active	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	17/02/2023	17/02/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	17/02/2023	17/02/2023
					CT3001	Implantable metal	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	17/02/2023	17/02/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	17/02/2023	17/02/2023
					CT3002	Implantable synthetic polymer	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	17/02/2023	17/02/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	17/02/2023	17/02/2023
					CT3006	Implantable inorganic material	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	17/02/2023	17/02/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	17/02/2023	17/02/2023
					CT314	Electrically powered	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	17/02/2023	17/02/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	17/02/2023	17/02/2023
					CT321	Long-term surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	17/02/2023	17/02/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	17/02/2023	17/02/2023
					CT449	Generators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	17/02/2023	17/02/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	17/02/2023	17/02/2023
					CT981	Single use	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	17/02/2023	17/02/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	17/02/2023	17/02/2023
					CT983	Surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	17/02/2023	17/02/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	17/02/2023	17/02/2023
12848	2023		47265	Dual-chamber implantable pacemaker, rate-responsive	CT141	Electrodes/Leads and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	04/04/2023	04/04/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	04/04/2023	04/04/2023
					CT145	Cardiovascular devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	04/04/2023	04/04/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	04/04/2023	04/04/2023
					CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	04/04/2023	04/04/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	04/04/2023	04/04/2023
					CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	04/04/2023	04/04/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	04/04/2023	04/04/2023
					CT233	Surgical	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	04/04/2023	04/04/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	04/04/2023	04/04/2023
					CT2406	Implantable	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	04/04/2023	04/04/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	04/04/2023	04/04/2023
					CT2407	Active implantable	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	04/04/2023	04/04/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	04/04/2023	04/04/2023
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	04/04/2023	04/04/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	04/04/2023	04/04/2023
					CT2727	Active implant devices/ancillaries	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	04/04/2023	04/04/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	04/04/2023	04/04/2023
					CT2750	Implantable cardiac electrical stimulators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	04/04/2023	04/04/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	04/04/2023	04/04/2023
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	04/04/2023	04/04/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	04/04/2023	04/04/2023
					CT2985	Active	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	04/04/2023	04/04/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	04/04/2023	04/04/2023
					CT3001	Implantable metal	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	04/04/2023	04/04/2023

							F1901	Additional Surgery	A24	No Apparent Device Problem	04/04/2023	04/04/2023
					CT3002	Implantable synthetic polymer	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	04/04/2023	04/04/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	04/04/2023	04/04/2023
					CT3006	Implantable inorganic material	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	04/04/2023	04/04/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	04/04/2023	04/04/2023
					CT314	Electrically powered	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	04/04/2023	04/04/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	04/04/2023	04/04/2023
					CT321	Long-term surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	04/04/2023	04/04/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	04/04/2023	04/04/2023
					CT449	Generators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	04/04/2023	04/04/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	04/04/2023	04/04/2023
					CT981	Single use	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	04/04/2023	04/04/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	04/04/2023	04/04/2023
					CT983	Surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	04/04/2023	04/04/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	04/04/2023	04/04/2023
12849	2023	All other reportable incidents	47263	Cardiac resynchronization therapy implantable pacemaker	CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A0908	Incorrect, Inadequate or Imprecise Result or Readings	17/03/2023	17/03/2023
							F08	Hospitalization or Prolonged Hospitalization	A0908	Incorrect, Inadequate or Imprecise Result or Readings	17/03/2023	17/03/2023
							F1905	Device Revision or Replacement	A0908	Incorrect, Inadequate or Imprecise Result or Readings	17/03/2023	17/03/2023
					CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A0908	Incorrect, Inadequate or Imprecise Result or Readings	17/03/2023	17/03/2023
							F08	Hospitalization or Prolonged Hospitalization	A0908	Incorrect, Inadequate or Imprecise Result or Readings	17/03/2023	17/03/2023
							F1905	Device Revision or Replacement	A0908	Incorrect, Inadequate or Imprecise Result or Readings	17/03/2023	17/03/2023
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A0908	Incorrect, Inadequate or Imprecise Result or Readings	17/03/2023	17/03/2023
							F08	Hospitalization or Prolonged Hospitalization	A0908	Incorrect, Inadequate or Imprecise Result or Readings	17/03/2023	17/03/2023
							F1905	Device Revision or Replacement	A0908	Incorrect, Inadequate or Imprecise Result or Readings	17/03/2023	17/03/2023
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A0908	Incorrect, Inadequate or Imprecise Result or Readings	17/03/2023	17/03/2023
							F08	Hospitalization or Prolonged Hospitalization	A0908	Incorrect, Inadequate or Imprecise Result or Readings	17/03/2023	17/03/2023
							F1905	Device Revision or Replacement	A0908	Incorrect, Inadequate or Imprecise Result or Readings	17/03/2023	17/03/2023
12850	2023	All other reportable incidents	47265	Dual-chamber implantable pacemaker, rate-responsive	CT141	Electrodes/Leads and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem		
							F1901	Additional Surgery	A24	No Apparent Device Problem		
					CT145	Cardiovascular devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem		
							F1901	Additional Surgery	A24	No Apparent Device Problem		
					CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem		
							F1901	Additional Surgery	A24	No Apparent Device Problem		
					CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem		
							F1901	Additional Surgery	A24	No Apparent Device Problem		
					CT233	Surgical	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem		
							F1901	Additional Surgery	A24	No Apparent Device Problem		
					CT2406	Implantable	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem		
							F1901	Additional Surgery	A24	No Apparent Device Problem		
					CT2407	Active implantable	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem		
							F1901	Additional Surgery	A24	No Apparent Device Problem		
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem		
							F1901	Additional Surgery	A24	No Apparent Device Problem		
					CT2727	Active implant devices/ancillaries	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem		
							F1901	Additional Surgery	A24	No Apparent Device Problem		
					CT2750	Implantable cardiac electrical stimulators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem		
							F1901	Additional Surgery	A24	No Apparent Device Problem		
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem		
							F1901	Additional Surgery	A24	No Apparent Device Problem		
					CT2985	Active	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem		
							F1901	Additional Surgery	A24	No Apparent Device Problem		
					CT3001	Implantable metal	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem		
							F1901	Additional Surgery	A24	No Apparent Device Problem		
					CT3002	Implantable synthetic polymer	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem		
							F1901	Additional Surgery	A24	No Apparent Device Problem		
					CT3006	Implantable inorganic material	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem		
							F1901	Additional Surgery	A24	No Apparent Device Problem		
					CT314	Electrically powered	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem		
							F1901	Additional Surgery	A24	No Apparent Device Problem		
					CT321	Long-term surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem		
							F1901	Additional Surgery	A24	No Apparent Device Problem		
					CT449	Generators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem		
							F1901	Additional Surgery	A24	No Apparent Device Problem		
					CT981	Single use	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem		
							F1901	Additional Surgery	A24	No Apparent Device Problem		
					CT983	Surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem		
							F1901	Additional Surgery	A24	No Apparent Device Problem		
12851	2023		47265	Dual-chamber implantable pacemaker, rate-responsive	CT141	Electrodes/Leads and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem		
							F1901	Additional Surgery	A24	No Apparent Device Problem		
					CT145	Cardiovascular devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem		
							F1901	Additional Surgery	A24	No Apparent Device Problem		
					CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem		
							F1901	Additional Surgery	A24	No Apparent Device Problem		
					CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem		
							F1901	Additional Surgery	A24	No Apparent Device Problem		
					CT233	Surgical	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem		
							F1901	Additional Surgery	A24	No Apparent Device Problem		
					CT2406	Implantable	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem		
							F1901	Additional Surgery	A24	No Apparent Device Problem		

					CT2407	Active implantable	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem		
							F1901	Additional Surgery	A24	No Apparent Device Problem		
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem		
							F1901	Additional Surgery	A24	No Apparent Device Problem		
					CT2727	Active implant devices/ancillaries	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem		
							F1901	Additional Surgery	A24	No Apparent Device Problem		
					CT2750	Implantable cardiac electrical stimulators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem		
							F1901	Additional Surgery	A24	No Apparent Device Problem		
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem		
							F1901	Additional Surgery	A24	No Apparent Device Problem		
					CT2985	Active	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem		
							F1901	Additional Surgery	A24	No Apparent Device Problem		
					CT3001	Implantable metal	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem		
							F1901	Additional Surgery	A24	No Apparent Device Problem		
					CT3002	Implantable synthetic polymer	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem		
							F1901	Additional Surgery	A24	No Apparent Device Problem		
					CT3006	Implantable inorganic material	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem		
							F1901	Additional Surgery	A24	No Apparent Device Problem		
					CT314	Electrically powered	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem		
							F1901	Additional Surgery	A24	No Apparent Device Problem		
					CT321	Long-term surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem		
							F1901	Additional Surgery	A24	No Apparent Device Problem		
					CT449	Generators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem		
							F1901	Additional Surgery	A24	No Apparent Device Problem		
					CT981	Single use	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem		
							F1901	Additional Surgery	A24	No Apparent Device Problem		
					CT983	Surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem		
							F1901	Additional Surgery	A24	No Apparent Device Problem		
12852	2023		47265	Dual-chamber implantable pacemaker, rate-responsive	CT141	Electrodes/Leads and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	04/04/2023	04/04/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	04/04/2023	04/04/2023
					CT145	Cardiovascular devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	04/04/2023	04/04/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	04/04/2023	04/04/2023
					CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	04/04/2023	04/04/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	04/04/2023	04/04/2023
					CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	04/04/2023	04/04/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	04/04/2023	04/04/2023
					CT233	Surgical	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	04/04/2023	04/04/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	04/04/2023	04/04/2023
					CT2406	Implantable	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	04/04/2023	04/04/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	04/04/2023	04/04/2023
					CT2407	Active implantable	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	04/04/2023	04/04/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	04/04/2023	04/04/2023
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	04/04/2023	04/04/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	04/04/2023	04/04/2023
					CT2727	Active implant devices/ancillaries	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	04/04/2023	04/04/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	04/04/2023	04/04/2023
					CT2750	Implantable cardiac electrical stimulators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	04/04/2023	04/04/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	04/04/2023	04/04/2023
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	04/04/2023	04/04/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	04/04/2023	04/04/2023
					CT2985	Active	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	04/04/2023	04/04/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	04/04/2023	04/04/2023
					CT3001	Implantable metal	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	04/04/2023	04/04/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	04/04/2023	04/04/2023
					CT3002	Implantable synthetic polymer	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	04/04/2023	04/04/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	04/04/2023	04/04/2023
					CT3006	Implantable inorganic material	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	04/04/2023	04/04/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	04/04/2023	04/04/2023
					CT314	Electrically powered	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	04/04/2023	04/04/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	04/04/2023	04/04/2023
					CT321	Long-term surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	04/04/2023	04/04/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	04/04/2023	04/04/2023
					CT449	Generators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	04/04/2023	04/04/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	04/04/2023	04/04/2023
					CT981	Single use	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	04/04/2023	04/04/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	04/04/2023	04/04/2023
					CT983	Surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	04/04/2023	04/04/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	04/04/2023	04/04/2023
12853	2023		47265	Dual-chamber implantable pacemaker, rate-responsive	CT141	Electrodes/Leads and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	04/04/2023	04/04/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	04/04/2023	04/04/2023
					CT145	Cardiovascular devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	04/04/2023	04/04/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	04/04/2023	04/04/2023
					CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	04/04/2023	04/04/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	04/04/2023	04/04/2023
					CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	04/04/2023	04/04/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	04/04/2023	04/04/2023
					CT233	Surgical	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	04/04/2023	04/04/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	04/04/2023	04/04/2023
					CT2406	Implantable	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	04/04/2023	04/04/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	04/04/2023	04/04/2023
					CT2407	Active implantable	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	04/04/2023	04/04/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	04/04/2023	04/04/2023
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	04/04/2023	04/04/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	04/04/2023	04/04/2023
					CT2727	Active implant devices/ancillaries	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	04/04/2023	04/04/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	04/04/2023	04/04/2023
					CT2750	Implantable cardiac electrical stimulators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	04/04/2023	04/04/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	04/04/2023	04/04/2023

					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	04/04/2023	04/04/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	04/04/2023	04/04/2023
					CT2985	Active	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	04/04/2023	04/04/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	04/04/2023	04/04/2023
					CT3001	Implantable metal	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	04/04/2023	04/04/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	04/04/2023	04/04/2023
					CT3002	Implantable synthetic polymer	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	04/04/2023	04/04/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	04/04/2023	04/04/2023
					CT3006	Implantable inorganic material	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	04/04/2023	04/04/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	04/04/2023	04/04/2023
					CT314	Electrically powered	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	04/04/2023	04/04/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	04/04/2023	04/04/2023
					CT321	Long-term surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	04/04/2023	04/04/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	04/04/2023	04/04/2023
					CT449	Generators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	04/04/2023	04/04/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	04/04/2023	04/04/2023
					CT981	Single use	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	04/04/2023	04/04/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	04/04/2023	04/04/2023
					CT983	Surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	04/04/2023	04/04/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	04/04/2023	04/04/2023
12854	2023	All other reportable incidents	17882	Physiologic-monitoring defibrillation system	CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070803	Failure to Power Up		
							F27	Problem Identified During Non-Clinical Procedure	A070803	Failure to Power Up		
12855	2023	All other reportable incidents	35852	Single-chamber implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
									A0908	Incorrect, Inadequate or Imprecise Result or Readings		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
									A0908	Incorrect, Inadequate or Imprecise Result or Readings		
12856	2023	All other reportable incidents	47265	Dual-chamber implantable pacemaker, rate-responsive	CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
12857	2023		47265	Dual-chamber implantable pacemaker, rate-responsive	CT141	Electrodes/Leads and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	29/03/2023	29/03/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	29/03/2023	29/03/2023
					CT145	Cardiovascular devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	29/03/2023	29/03/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	29/03/2023	29/03/2023
					CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	29/03/2023	29/03/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	29/03/2023	29/03/2023
					CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	29/03/2023	29/03/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	29/03/2023	29/03/2023
					CT233	Surgical	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	29/03/2023	29/03/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	29/03/2023	29/03/2023
					CT2406	Implantable	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	29/03/2023	29/03/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	29/03/2023	29/03/2023
					CT2407	Active implantable	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	29/03/2023	29/03/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	29/03/2023	29/03/2023
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	29/03/2023	29/03/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	29/03/2023	29/03/2023
					CT2727	Active implant devices/ancillaries	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	29/03/2023	29/03/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	29/03/2023	29/03/2023
					CT2750	Implantable cardiac electrical stimulators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	29/03/2023	29/03/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	29/03/2023	29/03/2023
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	29/03/2023	29/03/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	29/03/2023	29/03/2023
					CT2985	Active	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	29/03/2023	29/03/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	29/03/2023	29/03/2023
					CT3001	Implantable metal	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	29/03/2023	29/03/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	29/03/2023	29/03/2023
					CT3002	Implantable synthetic polymer	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	29/03/2023	29/03/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	29/03/2023	29/03/2023
					CT3006	Implantable inorganic material	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	29/03/2023	29/03/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	29/03/2023	29/03/2023
					CT314	Electrically powered	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	29/03/2023	29/03/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	29/03/2023	29/03/2023
					CT321	Long-term surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	29/03/2023	29/03/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	29/03/2023	29/03/2023
					CT449	Generators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	29/03/2023	29/03/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	29/03/2023	29/03/2023
					CT981	Single use	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	29/03/2023	29/03/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	29/03/2023	29/03/2023
					CT983	Surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	29/03/2023	29/03/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	29/03/2023	29/03/2023
12858	2023	All other reportable incidents	37265	Dual-chamber implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A072201	High Impedance		
							F26	No Health Consequences or Impact	A072201	High Impedance		
12859	2023	All other reportable incidents	35852	Single-chamber implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
12860	2023		47265	Dual-chamber implantable pacemaker, rate-responsive	CT141	Electrodes/Leads and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	06/04/2023	06/04/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	06/04/2023	06/04/2023
					CT145	Cardiovascular devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	06/04/2023	06/04/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	06/04/2023	06/04/2023

					CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	06/04/2023	06/04/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	06/04/2023	06/04/2023
					CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	06/04/2023	06/04/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	06/04/2023	06/04/2023
					CT233	Surgical	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	06/04/2023	06/04/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	06/04/2023	06/04/2023
					CT2406	Implantable	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	06/04/2023	06/04/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	06/04/2023	06/04/2023
					CT2407	Active implantable	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	06/04/2023	06/04/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	06/04/2023	06/04/2023
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	06/04/2023	06/04/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	06/04/2023	06/04/2023
					CT2727	Active implant devices/ancillaries	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	06/04/2023	06/04/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	06/04/2023	06/04/2023
					CT2750	Implantable cardiac electrical stimulators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	06/04/2023	06/04/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	06/04/2023	06/04/2023
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	06/04/2023	06/04/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	06/04/2023	06/04/2023
					CT2985	Active	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	06/04/2023	06/04/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	06/04/2023	06/04/2023
					CT3001	Implantable metal	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	06/04/2023	06/04/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	06/04/2023	06/04/2023
					CT3002	Implantable synthetic polymer	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	06/04/2023	06/04/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	06/04/2023	06/04/2023
					CT3006	Implantable inorganic material	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	06/04/2023	06/04/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	06/04/2023	06/04/2023
					CT314	Electrically powered	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	06/04/2023	06/04/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	06/04/2023	06/04/2023
					CT321	Long-term surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	06/04/2023	06/04/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	06/04/2023	06/04/2023
					CT449	Generators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	06/04/2023	06/04/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	06/04/2023	06/04/2023
					CT981	Single use	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	06/04/2023	06/04/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	06/04/2023	06/04/2023
					CT983	Surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	06/04/2023	06/04/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	06/04/2023	06/04/2023
12861	2023	All other reportable incidents	37265	Dual-chamber implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	04/04/2023	04/04/2023
							F1901	Additional Surgery	A070504	Premature Discharge of Battery	04/04/2023	04/04/2023
12862	2023	All other reportable incidents	47265	Dual-chamber implantable pacemaker, rate-responsive	CT141	Electrodes/Leads and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070101	Failure to Capture		
							F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A070101	Failure to Capture		
					CT145	Cardiovascular devices	E2403	No Clinical Signs, Symptoms or Conditions	A070101	Failure to Capture		
							F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A070101	Failure to Capture		
					CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070101	Failure to Capture		
							F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A070101	Failure to Capture		
					CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A070101	Failure to Capture		
							F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A070101	Failure to Capture		
					CT233	Surgical	E2403	No Clinical Signs, Symptoms or Conditions	A070101	Failure to Capture		
							F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A070101	Failure to Capture		
					CT2406	Implantable	E2403	No Clinical Signs, Symptoms or Conditions	A070101	Failure to Capture		
							F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A070101	Failure to Capture		
					CT2407	Active implantable	E2403	No Clinical Signs, Symptoms or Conditions	A070101	Failure to Capture		
							F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A070101	Failure to Capture		
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A070101	Failure to Capture		
							F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A070101	Failure to Capture		
					CT2727	Active implant devices/ancillaries	E2403	No Clinical Signs, Symptoms or Conditions	A070101	Failure to Capture		
							F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A070101	Failure to Capture		
					CT2750	Implantable cardiac electrical stimulators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070101	Failure to Capture		
							F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A070101	Failure to Capture		
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A070101	Failure to Capture		
							F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A070101	Failure to Capture		
					CT2985	Active	E2403	No Clinical Signs, Symptoms or Conditions	A070101	Failure to Capture		
							F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A070101	Failure to Capture		
					CT3001	Implantable metal	E2403	No Clinical Signs, Symptoms or Conditions	A070101	Failure to Capture		
							F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A070101	Failure to Capture		
					CT3002	Implantable synthetic polymer	E2403	No Clinical Signs, Symptoms or Conditions	A070101	Failure to Capture		
							F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A070101	Failure to Capture		
					CT3006	Implantable inorganic material	E2403	No Clinical Signs, Symptoms or Conditions	A070101	Failure to Capture		
							F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A070101	Failure to Capture		
					CT314	Electrically powered	E2403	No Clinical Signs, Symptoms or Conditions	A070101	Failure to Capture		
							F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A070101	Failure to Capture		
					CT321	Long-term surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A070101	Failure to Capture		
							F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A070101	Failure to Capture		
					CT449	Generators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070101	Failure to Capture		
							F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A070101	Failure to Capture		
					CT981	Single use	E2403	No Clinical Signs, Symptoms or Conditions	A070101	Failure to Capture		
							F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A070101	Failure to Capture		
					CT983	Surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A070101	Failure to Capture		
							F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A070101	Failure to Capture		
12863	2023	All other reportable incidents	17882	Physiologic-monitoring defibrillation system	CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070803	Failure to Power Up		
							F27	Problem Identified During Non-Clinical Procedure	A070803	Failure to Power Up		
12864	2023	All other reportable incidents	35852	Single-chamber implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	06/07/2023	06/07/2023
									A0908	Incorrect, Inadequate or Imprecise Result or Readings	06/07/2023	06/07/2023
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	06/07/2023	06/07/2023
									A0908	Incorrect, Inadequate or Imprecise Result or Readings	06/07/2023	06/07/2023
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	06/07/2023	06/07/2023

									A0908	Incorrect, Inadequate or Imprecise Result or Readings	06/07/2023	06/07/2023
12865	2023	All other reportable incidents	47265	Dual-chamber implantable pacemaker, rate-responsive	CT141	Electrodes/Leads and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT145	Cardiovascular devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT233	Surgical	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT2406	Implantable	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT2407	Active implantable	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT2727	Active implant devices/ancillaries	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT2750	Implantable cardiac electrical stimulators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT2985	Active	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT3001	Implantable metal	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT3002	Implantable synthetic polymer	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT3006	Implantable inorganic material	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT314	Electrically powered	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT321	Long-term surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT449	Generators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT981	Single use	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT983	Surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
12866	2023	All other reportable incidents	17882	Physiologic-monitoring defibrillation system	CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A071204	No Pacing		
							F27	Problem Identified During Non-Clinical Procedure	A071204	No Pacing		
12867	2023	All other reportable incidents	47940	Triple-chamber implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
									A0908	Incorrect, Inadequate or Imprecise Result or Readings		
							F24	Insufficient Health Impact Information	A070504	Premature Discharge of Battery		
									A0908	Incorrect, Inadequate or Imprecise Result or Readings		
12868	2023	All other reportable incidents	47265	Dual-chamber implantable pacemaker, rate-responsive	CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A1206	Misconnection		
							F08	Hospitalization or Prolonged Hospitalization	A1206	Misconnection		
							F12	Serious Injury/ Illness/ Impairment	A1206	Misconnection		
							F1905	Device Revision or Replacement	A1206	Misconnection		
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A1206	Misconnection		
							F08	Hospitalization or Prolonged Hospitalization	A1206	Misconnection		
							F12	Serious Injury/ Illness/ Impairment	A1206	Misconnection		
							F1905	Device Revision or Replacement	A1206	Misconnection		
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A1206	Misconnection		
							F08	Hospitalization or Prolonged Hospitalization	A1206	Misconnection		
							F12	Serious Injury/ Illness/ Impairment	A1206	Misconnection		
							F1905	Device Revision or Replacement	A1206	Misconnection		
12869	2023	All other reportable incidents	47263	Cardiac resynchronization therapy implantable pacemaker	CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A0908	Incorrect, Inadequate or Imprecise Result or Readings	06/04/2023	06/04/2023
							F08	Hospitalization or Prolonged Hospitalization	A0908	Incorrect, Inadequate or Imprecise Result or Readings	06/04/2023	06/04/2023
							F1905	Device Revision or Replacement	A0908	Incorrect, Inadequate or Imprecise Result or Readings	06/04/2023	06/04/2023
					CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A0908	Incorrect, Inadequate or Imprecise Result or Readings	06/04/2023	06/04/2023
							F08	Hospitalization or Prolonged Hospitalization	A0908	Incorrect, Inadequate or Imprecise Result or Readings	06/04/2023	06/04/2023
							F1905	Device Revision or Replacement	A0908	Incorrect, Inadequate or Imprecise Result or Readings	06/04/2023	06/04/2023
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A0908	Incorrect, Inadequate or Imprecise Result or Readings	06/04/2023	06/04/2023
							F08	Hospitalization or Prolonged Hospitalization	A0908	Incorrect, Inadequate or Imprecise Result or Readings	06/04/2023	06/04/2023
							F1905	Device Revision or Replacement	A0908	Incorrect, Inadequate or Imprecise Result or Readings	06/04/2023	06/04/2023
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A0908	Incorrect, Inadequate or Imprecise Result or Readings	06/04/2023	06/04/2023
							F08	Hospitalization or Prolonged Hospitalization	A0908	Incorrect, Inadequate or Imprecise Result or Readings	06/04/2023	06/04/2023
							F1905	Device Revision or Replacement	A0908	Incorrect, Inadequate or Imprecise Result or Readings	06/04/2023	06/04/2023
12870	2023	All other reportable incidents	35852	Single-chamber implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E2104	Electric Shock	A070909	Oversensing		
									A071302	Inappropriate/Inadequate Shock/Stimulation		
							F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A070909	Oversensing		
									A071302	Inappropriate/Inadequate Shock/Stimulation		
12871	2023	All other reportable incidents	37265	Dual-chamber implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E060109	Tachycardia	A0709	Device Sensing Problem		

								A071301	Failure to Deliver Shock/Stimulation		
						F05	Delay to Treatment/ Therapy	A0709	Device Sensing Problem		
								A071301	Failure to Deliver Shock/Stimulation		
						F1203	Life Threatening Illness or Injury	A0709	Device Sensing Problem		
								A071301	Failure to Deliver Shock/Stimulation		
12872	2023	All other reportable incidents	47265	Dual-chamber implantable pacemaker, rate-responsive	CT141	Electrodes/Leads and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
					CT145	Cardiovascular devices	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
					CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
					CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
					CT233	Surgical	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
					CT2406	Implantable	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
					CT2407	Active implantable	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
					CT2727	Active implant devices/ancillaries	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
					CT2750	Implantable cardiac electrical stimulators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
					CT2985	Active	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
					CT3001	Implantable metal	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
					CT3002	Implantable synthetic polymer	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
					CT3006	Implantable inorganic material	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
					CT314	Electrically powered	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
					CT321	Long-term surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
					CT449	Generators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
					CT981	Single use	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
					CT983	Surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
12873	2023	All other reportable incidents	17882	Physiologic-monitoring defibrillation system	CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A0719	Unexpected Shutdown		
							F27	Problem Identified During Non-Clinical Procedure	A0719	Unexpected Shutdown		
12874	2023	Death	17882	Physiologic-monitoring defibrillation system	CT1684	Cardiac pacemakers and associated devices	E0601	Arrhythmia	A23	Use of Device Problem		

							F02	Death	A23	Use of Device Problem		
12875	2023	All other reportable incidents	35223	Endocardial/interventricular septal pacing lead	CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A051201	Device Dislodged or Dislocated		
									A070101	Failure to Capture		
									A070909	Oversensing		
							F08	Hospitalization or Prolonged Hospitalization	A051201	Device Dislodged or Dislocated		
									A070101	Failure to Capture		
									A070909	Oversensing		
							F12	Serious Injury/ Illness/ Impairment	A051201	Device Dislodged or Dislocated		
									A070101	Failure to Capture		
									A070909	Oversensing		
							F1904	Device Repositioning	A051201	Device Dislodged or Dislocated		
									A070101	Failure to Capture		
									A070909	Oversensing		
12876	2023	All other reportable incidents			CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A0908	Incorrect, Inadequate or Imprecise Result or Readings	11/01/2023	11/01/2023
									A1107	Data Problem	11/01/2023	11/01/2023
							F08	Hospitalization or Prolonged Hospitalization	A0908	Incorrect, Inadequate or Imprecise Result or Readings	11/01/2023	11/01/2023
									A1107	Data Problem	11/01/2023	11/01/2023
							F1905	Device Revision or Replacement	A0908	Incorrect, Inadequate or Imprecise Result or Readings	11/01/2023	11/01/2023
									A1107	Data Problem	11/01/2023	11/01/2023
12877	2023	All other reportable incidents	37265	Dual-chamber implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E060109	Tachycardia	A0709	Device Sensing Problem		
									A071301	Failure to Deliver Shock/Stimulation		
							F1001	Absence of Treatment	A0709	Device Sensing Problem		
									A071301	Failure to Deliver Shock/Stimulation		
							F1203	Life Threatening Illness or Injury	A0709	Device Sensing Problem		
									A071301	Failure to Deliver Shock/Stimulation		
12878	2023	All other reportable incidents	36241	Pacemaker/defibrillator lead	CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A072202	Low Impedance		
									A090801	Signal Artifact/Noise		
							F26	No Health Consequences or Impact	A072202	Low Impedance		
									A090801	Signal Artifact/Noise		
12879	2023	All other reportable incidents	47265	Dual-chamber implantable pacemaker, rate-responsive	CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A0908	Incorrect, Inadequate or Imprecise Result or Readings	29/03/2023	29/03/2023
							F08	Hospitalization or Prolonged Hospitalization	A0908	Incorrect, Inadequate or Imprecise Result or Readings	29/03/2023	29/03/2023
							F1905	Device Revision or Replacement	A0908	Incorrect, Inadequate or Imprecise Result or Readings	29/03/2023	29/03/2023
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A0908	Incorrect, Inadequate or Imprecise Result or Readings	29/03/2023	29/03/2023
							F08	Hospitalization or Prolonged Hospitalization	A0908	Incorrect, Inadequate or Imprecise Result or Readings	29/03/2023	29/03/2023
							F1905	Device Revision or Replacement	A0908	Incorrect, Inadequate or Imprecise Result or Readings	29/03/2023	29/03/2023
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A0908	Incorrect, Inadequate or Imprecise Result or Readings	29/03/2023	29/03/2023
							F08	Hospitalization or Prolonged Hospitalization	A0908	Incorrect, Inadequate or Imprecise Result or Readings	29/03/2023	29/03/2023
							F1905	Device Revision or Replacement	A0908	Incorrect, Inadequate or Imprecise Result or Readings	29/03/2023	29/03/2023
12880	2023	All other reportable incidents	47265	Dual-chamber implantable pacemaker, rate-responsive	CT141	Electrodes/Leads and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
							F24	Insufficient Health Impact Information	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
					CT145	Cardiovascular devices	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
							F24	Insufficient Health Impact Information	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
					CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
							F24	Insufficient Health Impact Information	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
					CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
							F24	Insufficient Health Impact Information	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
					CT233	Surgical	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
							F24	Insufficient Health Impact Information	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
					CT2406	Implantable	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
							F24	Insufficient Health Impact Information	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
					CT2407	Active implantable	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
							F24	Insufficient Health Impact Information	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
							F24	Insufficient Health Impact Information	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
					CT2727	Active implant devices/ancillaries	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
							F24	Insufficient Health Impact Information	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
					CT2750	Implantable cardiac electrical stimulators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
							F24	Insufficient Health Impact Information	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem		

									A070504	Premature Discharge of Battery		
						F24	Insufficient Health Impact Information	A0705	Battery Problem		
								A070504	Premature Discharge of Battery		
					CT2985	Active	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
							F24	Insufficient Health Impact Information	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
					CT3001	Implantable metal	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
							F24	Insufficient Health Impact Information	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
					CT3002	Implantable synthetic polymer	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
							F24	Insufficient Health Impact Information	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
					CT3006	Implantable inorganic material	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
							F24	Insufficient Health Impact Information	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
					CT314	Electrically powered	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
							F24	Insufficient Health Impact Information	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
					CT321	Long-term surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
							F24	Insufficient Health Impact Information	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
					CT449	Generators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
							F24	Insufficient Health Impact Information	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
					CT981	Single use	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
							F24	Insufficient Health Impact Information	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
					CT983	Surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
							F24	Insufficient Health Impact Information	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
12881	2023		47265	Dual-chamber implantable pacemaker, rate-responsive	CT141	Electrodes/Leads and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	31/03/2023	31/03/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	31/03/2023	31/03/2023
					CT145	Cardiovascular devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	31/03/2023	31/03/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	31/03/2023	31/03/2023
					CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	31/03/2023	31/03/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	31/03/2023	31/03/2023
					CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	31/03/2023	31/03/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	31/03/2023	31/03/2023
					CT233	Surgical	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	31/03/2023	31/03/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	31/03/2023	31/03/2023
					CT2406	Implantable	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	31/03/2023	31/03/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	31/03/2023	31/03/2023
					CT2407	Active implantable	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	31/03/2023	31/03/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	31/03/2023	31/03/2023
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	31/03/2023	31/03/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	31/03/2023	31/03/2023
					CT2727	Active implant devices/ancillaries	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	31/03/2023	31/03/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	31/03/2023	31/03/2023
					CT2750	Implantable cardiac electrical stimulators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	31/03/2023	31/03/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	31/03/2023	31/03/2023
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	31/03/2023	31/03/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	31/03/2023	31/03/2023
					CT2985	Active	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	31/03/2023	31/03/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	31/03/2023	31/03/2023
					CT3001	Implantable metal	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	31/03/2023	31/03/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	31/03/2023	31/03/2023
					CT3002	Implantable synthetic polymer	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	31/03/2023	31/03/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	31/03/2023	31/03/2023
					CT3006	Implantable inorganic material	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	31/03/2023	31/03/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	31/03/2023	31/03/2023
					CT314	Electrically powered	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	31/03/2023	31/03/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	31/03/2023	31/03/2023
					CT321	Long-term surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	31/03/2023	31/03/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	31/03/2023	31/03/2023
					CT449	Generators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	31/03/2023	31/03/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	31/03/2023	31/03/2023
					CT981	Single use	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	31/03/2023	31/03/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	31/03/2023	31/03/2023
					CT983	Surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	31/03/2023	31/03/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	31/03/2023	31/03/2023
12882	2023	All other reportable incidents	37265	Dual-chamber implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A07	Electrical /Electronic Property Problem		
							F26	No Health Consequences or Impact	A07	Electrical /Electronic Property Problem		
12883	2023	All other reportable incidents	35223	Endocardial/interventricular septal pacing lead	CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A051201	Device Dislodged or Dislocated		
									A070909	Oversensing		
							F08	Hospitalization or Prolonged Hospitalization	A051201	Device Dislodged or Dislocated		
									A070909	Oversensing		
							F12	Serious Injury/ Illness/ Impairment	A051201	Device Dislodged or Dislocated		
									A070909	Oversensing		
							F1904	Device Repositioning	A051201	Device Dislodged or Dislocated		

									A070909	Oversensing		
12884	2023		12913	Pacemaker, implantable	CT145	Cardiovascular devices	E2401	Insufficient Information	A01	Patient Device Interaction Problem		
							F19	Surgical Intervention	A01	Patient Device Interaction Problem		
							F26	No Health Consequences or Impact	A01	Patient Device Interaction Problem		
					CT1684	Cardiac pacemakers and associated devices	E2401	Insufficient Information	A01	Patient Device Interaction Problem		
							F19	Surgical Intervention	A01	Patient Device Interaction Problem		
							F26	No Health Consequences or Impact	A01	Patient Device Interaction Problem		
					CT1685	Cardiac pacemakers	E2401	Insufficient Information	A01	Patient Device Interaction Problem		
							F19	Surgical Intervention	A01	Patient Device Interaction Problem		
							F26	No Health Consequences or Impact	A01	Patient Device Interaction Problem		
					CT2406	Implantable	E2401	Insufficient Information	A01	Patient Device Interaction Problem		
							F19	Surgical Intervention	A01	Patient Device Interaction Problem		
							F26	No Health Consequences or Impact	A01	Patient Device Interaction Problem		
					CT2407	Active implantable	E2401	Insufficient Information	A01	Patient Device Interaction Problem		
							F19	Surgical Intervention	A01	Patient Device Interaction Problem		
							F26	No Health Consequences or Impact	A01	Patient Device Interaction Problem		
					CT257	Implantable cardiac pacemakers	E2401	Insufficient Information	A01	Patient Device Interaction Problem		
							F19	Surgical Intervention	A01	Patient Device Interaction Problem		
							F26	No Health Consequences or Impact	A01	Patient Device Interaction Problem		
					CT2727	Active implant devices/ancillaries	E2401	Insufficient Information	A01	Patient Device Interaction Problem		
							F19	Surgical Intervention	A01	Patient Device Interaction Problem		
							F26	No Health Consequences or Impact	A01	Patient Device Interaction Problem		
					CT2750	Implantable cardiac electrical stimulators and associated devices	E2401	Insufficient Information	A01	Patient Device Interaction Problem		
							F19	Surgical Intervention	A01	Patient Device Interaction Problem		
							F26	No Health Consequences or Impact	A01	Patient Device Interaction Problem		
					CT2866	Implantable cardiac electrical stimulators	E2401	Insufficient Information	A01	Patient Device Interaction Problem		
							F19	Surgical Intervention	A01	Patient Device Interaction Problem		
							F26	No Health Consequences or Impact	A01	Patient Device Interaction Problem		
					CT2985	Active	E2401	Insufficient Information	A01	Patient Device Interaction Problem		
							F19	Surgical Intervention	A01	Patient Device Interaction Problem		
							F26	No Health Consequences or Impact	A01	Patient Device Interaction Problem		
					CT314	Electrically powered	E2401	Insufficient Information	A01	Patient Device Interaction Problem		
							F19	Surgical Intervention	A01	Patient Device Interaction Problem		
							F26	No Health Consequences or Impact	A01	Patient Device Interaction Problem		
12885	2023	All other reportable incidents	35852	Single-chamber implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E2104	Electric Shock	A070909	Oversensing		
									A070910	Undersensing		
									A071302	Inappropriate/Inadequate Shock/Stimulation		
							F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A070909	Oversensing		
									A070910	Undersensing		
									A071302	Inappropriate/Inadequate Shock/Stimulation		
12886	2023	All other reportable incidents	47265	Dual-chamber implantable pacemaker, rate-responsive	CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
12887	2023	All other reportable incidents	47265	Dual-chamber implantable pacemaker, rate-responsive	CT141	Electrodes/Leads and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT145	Cardiovascular devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT233	Surgical	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT2406	Implantable	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT2407	Active implantable	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT2727	Active implant devices/ancillaries	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT2750	Implantable cardiac electrical stimulators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT2985	Active	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT3001	Implantable metal	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT3002	Implantable synthetic polymer	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT3006	Implantable inorganic material	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT314	Electrically powered	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT321	Long-term surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT449	Generators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT981	Single use	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT983	Surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
12888	2023	All other reportable incidents	17882	Physiologic-monitoring defibrillation system	CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A0719	Unexpected Shutdown		
							F27	Problem Identified During Non-Clinical Procedure	A0719	Unexpected Shutdown		

12889	2023	All other reportable incidents	47265	Dual-chamber implantable pacemaker, rate-responsive	CT141	Electrodes/Leads and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A071102	Failure to Interrogate	06/04/2023	06/04/2023
							F1901	Additional Surgery	A071102	Failure to Interrogate	06/04/2023	06/04/2023
					CT145	Cardiovascular devices	E2403	No Clinical Signs, Symptoms or Conditions	A071102	Failure to Interrogate	06/04/2023	06/04/2023
							F1901	Additional Surgery	A071102	Failure to Interrogate	06/04/2023	06/04/2023
					CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A071102	Failure to Interrogate	06/04/2023	06/04/2023
							F1901	Additional Surgery	A071102	Failure to Interrogate	06/04/2023	06/04/2023
					CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A071102	Failure to Interrogate	06/04/2023	06/04/2023
							F1901	Additional Surgery	A071102	Failure to Interrogate	06/04/2023	06/04/2023
					CT233	Surgical	E2403	No Clinical Signs, Symptoms or Conditions	A071102	Failure to Interrogate	06/04/2023	06/04/2023
							F1901	Additional Surgery	A071102	Failure to Interrogate	06/04/2023	06/04/2023
					CT2406	Implantable	E2403	No Clinical Signs, Symptoms or Conditions	A071102	Failure to Interrogate	06/04/2023	06/04/2023
							F1901	Additional Surgery	A071102	Failure to Interrogate	06/04/2023	06/04/2023
					CT2407	Active implantable	E2403	No Clinical Signs, Symptoms or Conditions	A071102	Failure to Interrogate	06/04/2023	06/04/2023
							F1901	Additional Surgery	A071102	Failure to Interrogate	06/04/2023	06/04/2023
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A071102	Failure to Interrogate	06/04/2023	06/04/2023
							F1901	Additional Surgery	A071102	Failure to Interrogate	06/04/2023	06/04/2023
					CT2727	Active implant devices/ancillaries	E2403	No Clinical Signs, Symptoms or Conditions	A071102	Failure to Interrogate	06/04/2023	06/04/2023
							F1901	Additional Surgery	A071102	Failure to Interrogate	06/04/2023	06/04/2023
					CT2750	Implantable cardiac electrical stimulators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A071102	Failure to Interrogate	06/04/2023	06/04/2023
							F1901	Additional Surgery	A071102	Failure to Interrogate	06/04/2023	06/04/2023
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A071102	Failure to Interrogate	06/04/2023	06/04/2023
							F1901	Additional Surgery	A071102	Failure to Interrogate	06/04/2023	06/04/2023
					CT2985	Active	E2403	No Clinical Signs, Symptoms or Conditions	A071102	Failure to Interrogate	06/04/2023	06/04/2023
							F1901	Additional Surgery	A071102	Failure to Interrogate	06/04/2023	06/04/2023
					CT3001	Implantable metal	E2403	No Clinical Signs, Symptoms or Conditions	A071102	Failure to Interrogate	06/04/2023	06/04/2023
							F1901	Additional Surgery	A071102	Failure to Interrogate	06/04/2023	06/04/2023
					CT3002	Implantable synthetic polymer	E2403	No Clinical Signs, Symptoms or Conditions	A071102	Failure to Interrogate	06/04/2023	06/04/2023
							F1901	Additional Surgery	A071102	Failure to Interrogate	06/04/2023	06/04/2023
					CT3006	Implantable inorganic material	E2403	No Clinical Signs, Symptoms or Conditions	A071102	Failure to Interrogate	06/04/2023	06/04/2023
							F1901	Additional Surgery	A071102	Failure to Interrogate	06/04/2023	06/04/2023
					CT314	Electrically powered	E2403	No Clinical Signs, Symptoms or Conditions	A071102	Failure to Interrogate	06/04/2023	06/04/2023
							F1901	Additional Surgery	A071102	Failure to Interrogate	06/04/2023	06/04/2023
					CT321	Long-term surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A071102	Failure to Interrogate	06/04/2023	06/04/2023
							F1901	Additional Surgery	A071102	Failure to Interrogate	06/04/2023	06/04/2023
					CT449	Generators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A071102	Failure to Interrogate	06/04/2023	06/04/2023
							F1901	Additional Surgery	A071102	Failure to Interrogate	06/04/2023	06/04/2023
					CT981	Single use	E2403	No Clinical Signs, Symptoms or Conditions	A071102	Failure to Interrogate	06/04/2023	06/04/2023
							F1901	Additional Surgery	A071102	Failure to Interrogate	06/04/2023	06/04/2023
					CT983	Surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A071102	Failure to Interrogate	06/04/2023	06/04/2023
							F1901	Additional Surgery	A071102	Failure to Interrogate	06/04/2023	06/04/2023
12890	2023	All other reportable incidents	47265	Dual-chamber implantable pacemaker, rate-responsive	CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
12891	2023	All other reportable incidents	47263	Cardiac resynchronization therapy implantable pacemaker	CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	31/08/2022	31/08/2022
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	31/08/2022	31/08/2022
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	31/08/2022	31/08/2022
					CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	31/08/2022	31/08/2022
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	31/08/2022	31/08/2022
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	31/08/2022	31/08/2022
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	31/08/2022	31/08/2022
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	31/08/2022	31/08/2022
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	31/08/2022	31/08/2022
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	31/08/2022	31/08/2022
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	31/08/2022	31/08/2022
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	31/08/2022	31/08/2022
12892	2023	All other reportable incidents	35852	Single-chamber implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A0709	Device Sensing Problem		
									A070909	Oversensing		
									A090801	Signal Artifact/Noise		
									A1301	Failure to Read Input Signal		
							F26	No Health Consequences or Impact	A0709	Device Sensing Problem		
									A070909	Oversensing		
									A090801	Signal Artifact/Noise		
									A1301	Failure to Read Input Signal		
12893	2023	All other reportable incidents	37265	Dual-chamber implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A070909	Oversensing		
									A090801	Signal Artifact/Noise		
							F26	No Health Consequences or Impact	A070909	Oversensing		
									A090801	Signal Artifact/Noise		
12894	2023	All other reportable incidents	35223	Endocardial/interventricular septal pacing lead	CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A072201	High Impedance		
							F26	No Health Consequences or Impact	A072201	High Impedance		
12895	2023	All other reportable incidents	35223	Endocardial/interventricular septal pacing lead	CT1684	Cardiac pacemakers and associated devices	E0109	Convulsion/Seizure	A070104	Unstable Capture Threshold		
									A072201	High Impedance		
							F08	Hospitalization or Prolonged Hospitalization	A070104	Unstable Capture Threshold		
									A072201	High Impedance		
							F12	Serious Injury/ Illness/ Impairment	A070104	Unstable Capture Threshold		
									A072201	High Impedance		
12896	2023	All other reportable incidents	47265	Dual-chamber implantable pacemaker, rate-responsive			E2403	No Clinical Signs, Symptoms or Conditions	A072202	Low Impedance	16/05/2025	16/05/2025
							F26	No Health Consequences or Impact	A072202	Low Impedance	16/05/2025	16/05/2025
12897	2023	All other reportable incidents	47265	Dual-chamber implantable pacemaker, rate-responsive	CT141	Electrodes/Leads and associated devices	E060106	Heart Block	A0712	Pacing Problem		
									A160105	False Alarm		

						F08	Hospitalization or Prolonged Hospitalization	A0712	Pacing Problem		
								A160105	False Alarm		
						F12	Serious Injury/ Illness/ Impairment	A0712	Pacing Problem		
								A160105	False Alarm		
						F19	Surgical Intervention	A0712	Pacing Problem		
								A160105	False Alarm		
						F23	Unexpected Medical Intervention	A0712	Pacing Problem		
								A160105	False Alarm		
					CT145	Cardiovascular devices	E060106	Heart Block	A0712	Pacing Problem		
									A160105	False Alarm		
							F08	Hospitalization or Prolonged Hospitalization	A0712	Pacing Problem		
									A160105	False Alarm		
							F12	Serious Injury/ Illness/ Impairment	A0712	Pacing Problem		
									A160105	False Alarm		
							F19	Surgical Intervention	A0712	Pacing Problem		
									A160105	False Alarm		
							F23	Unexpected Medical Intervention	A0712	Pacing Problem		
									A160105	False Alarm		
					CT1684	Cardiac pacemakers and associated devices	E060106	Heart Block	A0712	Pacing Problem		
									A160105	False Alarm		
							F08	Hospitalization or Prolonged Hospitalization	A0712	Pacing Problem		
									A160105	False Alarm		
							F12	Serious Injury/ Illness/ Impairment	A0712	Pacing Problem		
									A160105	False Alarm		
							F19	Surgical Intervention	A0712	Pacing Problem		
									A160105	False Alarm		
							F23	Unexpected Medical Intervention	A0712	Pacing Problem		
									A160105	False Alarm		
					CT1685	Cardiac pacemakers	E060106	Heart Block	A0712	Pacing Problem		
									A160105	False Alarm		
							F08	Hospitalization or Prolonged Hospitalization	A0712	Pacing Problem		
									A160105	False Alarm		
							F12	Serious Injury/ Illness/ Impairment	A0712	Pacing Problem		
									A160105	False Alarm		
							F19	Surgical Intervention	A0712	Pacing Problem		
									A160105	False Alarm		
							F23	Unexpected Medical Intervention	A0712	Pacing Problem		
									A160105	False Alarm		
					CT233	Surgical	E060106	Heart Block	A0712	Pacing Problem		
									A160105	False Alarm		
							F08	Hospitalization or Prolonged Hospitalization	A0712	Pacing Problem		
									A160105	False Alarm		
							F12	Serious Injury/ Illness/ Impairment	A0712	Pacing Problem		
									A160105	False Alarm		
							F19	Surgical Intervention	A0712	Pacing Problem		
									A160105	False Alarm		
							F23	Unexpected Medical Intervention	A0712	Pacing Problem		
									A160105	False Alarm		
					CT2406	Implantable	E060106	Heart Block	A0712	Pacing Problem		
									A160105	False Alarm		
							F08	Hospitalization or Prolonged Hospitalization	A0712	Pacing Problem		
									A160105	False Alarm		
							F12	Serious Injury/ Illness/ Impairment	A0712	Pacing Problem		
									A160105	False Alarm		
							F19	Surgical Intervention	A0712	Pacing Problem		
									A160105	False Alarm		
							F23	Unexpected Medical Intervention	A0712	Pacing Problem		
									A160105	False Alarm		
					CT2407	Active implantable	E060106	Heart Block	A0712	Pacing Problem		
									A160105	False Alarm		
							F08	Hospitalization or Prolonged Hospitalization	A0712	Pacing Problem		
									A160105	False Alarm		
							F12	Serious Injury/ Illness/ Impairment	A0712	Pacing Problem		
									A160105	False Alarm		
							F19	Surgical Intervention	A0712	Pacing Problem		
									A160105	False Alarm		
							F23	Unexpected Medical Intervention	A0712	Pacing Problem		
									A160105	False Alarm		
					CT257	Implantable cardiac pacemakers	E060106	Heart Block	A0712	Pacing Problem		
									A160105	False Alarm		
							F08	Hospitalization or Prolonged Hospitalization	A0712	Pacing Problem		
									A160105	False Alarm		
							F12	Serious Injury/ Illness/ Impairment	A0712	Pacing Problem		
									A160105	False Alarm		
							F19	Surgical Intervention	A0712	Pacing Problem		
									A160105	False Alarm		
							F23	Unexpected Medical Intervention	A0712	Pacing Problem		
									A160105	False Alarm		
					CT2727	Active implant devices/ancillaries	E060106	Heart Block	A0712	Pacing Problem		
									A160105	False Alarm		
							F08	Hospitalization or Prolonged Hospitalization	A0712	Pacing Problem		
									A160105	False Alarm		
							F12	Serious Injury/ Illness/ Impairment	A0712	Pacing Problem		
									A160105	False Alarm		
							F19	Surgical Intervention	A0712	Pacing Problem		
									A160105	False Alarm		
							F23	Unexpected Medical Intervention	A0712	Pacing Problem		
									A160105	False Alarm		

					CT2750	Implantable cardiac electrical stimulators and associated devices	E060106	Heart Block	A0712	Pacing Problem		
									A160105	False Alarm		
							F08	Hospitalization or Prolonged Hospitalization	A0712	Pacing Problem		
									A160105	False Alarm		
							F12	Serious Injury/ Illness/ Impairment	A0712	Pacing Problem		
									A160105	False Alarm		
							F19	Surgical Intervention	A0712	Pacing Problem		
									A160105	False Alarm		
							F23	Unexpected Medical Intervention	A0712	Pacing Problem		
									A160105	False Alarm		
					CT2866	Implantable cardiac electrical stimulators	E060106	Heart Block	A0712	Pacing Problem		
									A160105	False Alarm		
							F08	Hospitalization or Prolonged Hospitalization	A0712	Pacing Problem		
									A160105	False Alarm		
							F12	Serious Injury/ Illness/ Impairment	A0712	Pacing Problem		
									A160105	False Alarm		
							F19	Surgical Intervention	A0712	Pacing Problem		
									A160105	False Alarm		
							F23	Unexpected Medical Intervention	A0712	Pacing Problem		
									A160105	False Alarm		
					CT2985	Active	E060106	Heart Block	A0712	Pacing Problem		
									A160105	False Alarm		
							F08	Hospitalization or Prolonged Hospitalization	A0712	Pacing Problem		
									A160105	False Alarm		
							F12	Serious Injury/ Illness/ Impairment	A0712	Pacing Problem		
									A160105	False Alarm		
							F19	Surgical Intervention	A0712	Pacing Problem		
									A160105	False Alarm		
							F23	Unexpected Medical Intervention	A0712	Pacing Problem		
									A160105	False Alarm		
					CT3001	Implantable metal	E060106	Heart Block	A0712	Pacing Problem		
									A160105	False Alarm		
							F08	Hospitalization or Prolonged Hospitalization	A0712	Pacing Problem		
									A160105	False Alarm		
							F12	Serious Injury/ Illness/ Impairment	A0712	Pacing Problem		
									A160105	False Alarm		
							F19	Surgical Intervention	A0712	Pacing Problem		
									A160105	False Alarm		
							F23	Unexpected Medical Intervention	A0712	Pacing Problem		
									A160105	False Alarm		
					CT3002	Implantable synthetic polymer	E060106	Heart Block	A0712	Pacing Problem		
									A160105	False Alarm		
							F08	Hospitalization or Prolonged Hospitalization	A0712	Pacing Problem		
									A160105	False Alarm		
							F12	Serious Injury/ Illness/ Impairment	A0712	Pacing Problem		
									A160105	False Alarm		
							F19	Surgical Intervention	A0712	Pacing Problem		
									A160105	False Alarm		
							F23	Unexpected Medical Intervention	A0712	Pacing Problem		
									A160105	False Alarm		
					CT3006	Implantable inorganic material	E060106	Heart Block	A0712	Pacing Problem		
									A160105	False Alarm		
							F08	Hospitalization or Prolonged Hospitalization	A0712	Pacing Problem		
									A160105	False Alarm		
							F12	Serious Injury/ Illness/ Impairment	A0712	Pacing Problem		
									A160105	False Alarm		
							F19	Surgical Intervention	A0712	Pacing Problem		
									A160105	False Alarm		
							F23	Unexpected Medical Intervention	A0712	Pacing Problem		
									A160105	False Alarm		
					CT314	Electrically powered	E060106	Heart Block	A0712	Pacing Problem		
									A160105	False Alarm		
							F08	Hospitalization or Prolonged Hospitalization	A0712	Pacing Problem		
									A160105	False Alarm		
							F12	Serious Injury/ Illness/ Impairment	A0712	Pacing Problem		
									A160105	False Alarm		
							F19	Surgical Intervention	A0712	Pacing Problem		
									A160105	False Alarm		
							F23	Unexpected Medical Intervention	A0712	Pacing Problem		
									A160105	False Alarm		
					CT321	Long-term surgical invasive	E060106	Heart Block	A0712	Pacing Problem		
									A160105	False Alarm		
							F08	Hospitalization or Prolonged Hospitalization	A0712	Pacing Problem		
									A160105	False Alarm		
							F12	Serious Injury/ Illness/ Impairment	A0712	Pacing Problem		
									A160105	False Alarm		
							F19	Surgical Intervention	A0712	Pacing Problem		
									A160105	False Alarm		
							F23	Unexpected Medical Intervention	A0712	Pacing Problem		
									A160105	False Alarm		
					CT449	Generators and associated devices	E060106	Heart Block	A0712	Pacing Problem		
									A160105	False Alarm		
							F08	Hospitalization or Prolonged Hospitalization	A0712	Pacing Problem		
									A160105	False Alarm		
							F12	Serious Injury/ Illness/ Impairment	A0712	Pacing Problem		
									A160105	False Alarm		
							F19	Surgical Intervention	A0712	Pacing Problem		
									A160105	False Alarm		

							F23	Unexpected Medical Intervention	A0712	Pacing Problem		
									A160105	False Alarm		
					CT981	Single use	E060106	Heart Block	A0712	Pacing Problem		
									A160105	False Alarm		
							F08	Hospitalization or Prolonged Hospitalization	A0712	Pacing Problem		
									A160105	False Alarm		
							F12	Serious Injury/ Illness/ Impairment	A0712	Pacing Problem		
									A160105	False Alarm		
							F19	Surgical Intervention	A0712	Pacing Problem		
									A160105	False Alarm		
							F23	Unexpected Medical Intervention	A0712	Pacing Problem		
									A160105	False Alarm		
					CT983	Surgical invasive	E060106	Heart Block	A0712	Pacing Problem		
									A160105	False Alarm		
							F08	Hospitalization or Prolonged Hospitalization	A0712	Pacing Problem		
									A160105	False Alarm		
							F12	Serious Injury/ Illness/ Impairment	A0712	Pacing Problem		
									A160105	False Alarm		
							F19	Surgical Intervention	A0712	Pacing Problem		
									A160105	False Alarm		
							F23	Unexpected Medical Intervention	A0712	Pacing Problem		
									A160105	False Alarm		
12898	2023	All other reportable incidents	35223	Endocardial/interventricular septal pacing lead	CT1684	Cardiac pacemakers and associated devices	E0604	Cardiac Perforation	A24	No Apparent Device Problem		
							E0731	Pleural Effusion	A24	No Apparent Device Problem		
							E0734	Pneumothorax	A24	No Apparent Device Problem		
							F08	Hospitalization or Prolonged Hospitalization	A24	No Apparent Device Problem		
							F1203	Life Threatening Illness or Injury	A24	No Apparent Device Problem		
							F1904	Device Repositioning	A24	No Apparent Device Problem		
12899	2023	All other reportable incidents	35223	Endocardial/interventricular septal pacing lead	CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070909	Oversensing		
							F26	No Health Consequences or Impact	A070909	Oversensing		
12900	2023	All other reportable incidents	35223	Endocardial/interventricular septal pacing lead	CT1684	Cardiac pacemakers and associated devices	E2103	Device Overstimulation of Tissue	A07	Electrical /Electronic Property Problem		
									A070102	High Capture Threshold		
									A070909	Oversensing		
									A072201	High Impedance		
									A090801	Signal Artifact/Noise		
							F23	Unexpected Medical Intervention	A07	Electrical /Electronic Property Problem		
									A070102	High Capture Threshold		
									A070909	Oversensing		
									A072201	High Impedance		
									A090801	Signal Artifact/Noise		
12901	2023	All other reportable incidents	35223	Endocardial/interventricular septal pacing lead	CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A040101	Fracture		
									A070909	Oversensing		
							F26	No Health Consequences or Impact	A040101	Fracture		
									A070909	Oversensing		
12902	2023	All other reportable incidents	35852	Single-chamber implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A070909	Oversensing	03/02/2023	03/02/2023
									A180103	Device Contamination with Body Fluid	03/02/2023	03/02/2023
									A180104	Device Contamination with Chemical or Other Material	03/02/2023	03/02/2023
							F08	Hospitalization or Prolonged Hospitalization	A070909	Oversensing	03/02/2023	03/02/2023
									A180103	Device Contamination with Body Fluid	03/02/2023	03/02/2023
									A180104	Device Contamination with Chemical or Other Material	03/02/2023	03/02/2023
							F1905	Device Revision or Replacement	A070909	Oversensing	03/02/2023	03/02/2023
									A180103	Device Contamination with Body Fluid	03/02/2023	03/02/2023
									A180104	Device Contamination with Chemical or Other Material	03/02/2023	03/02/2023
12903	2023	All other reportable incidents	35852	Single-chamber implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E2330	Pain	A010402	Migration	25/01/2022	25/01/2022
							F08	Hospitalization or Prolonged Hospitalization	A010402	Migration	25/01/2022	25/01/2022
							F1905	Device Revision or Replacement	A010402	Migration	25/01/2022	25/01/2022
12904	2023	All other reportable incidents	47263	Cardiac resynchronization therapy implantable pacemaker	CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A0908	Incorrect, Inadequate or Imprecise Result or Readings	12/04/2023	12/04/2023
							F08	Hospitalization or Prolonged Hospitalization	A0908	Incorrect, Inadequate or Imprecise Result or Readings	12/04/2023	12/04/2023
							F1905	Device Revision or Replacement	A0908	Incorrect, Inadequate or Imprecise Result or Readings	12/04/2023	12/04/2023
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A0908	Incorrect, Inadequate or Imprecise Result or Readings	12/04/2023	12/04/2023
							F08	Hospitalization or Prolonged Hospitalization	A0908	Incorrect, Inadequate or Imprecise Result or Readings	12/04/2023	12/04/2023
							F1905	Device Revision or Replacement	A0908	Incorrect, Inadequate or Imprecise Result or Readings	12/04/2023	12/04/2023
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A0908	Incorrect, Inadequate or Imprecise Result or Readings	12/04/2023	12/04/2023
							F08	Hospitalization or Prolonged Hospitalization	A0908	Incorrect, Inadequate or Imprecise Result or Readings	12/04/2023	12/04/2023
							F1905	Device Revision or Replacement	A0908	Incorrect, Inadequate or Imprecise Result or Readings	12/04/2023	12/04/2023
12905	2023	All other reportable incidents	35223	Endocardial/interventricular septal pacing lead	CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A07	Electrical /Electronic Property Problem		
									A072201	High Impedance		
							F26	No Health Consequences or Impact	A07	Electrical /Electronic Property Problem		
									A072201	High Impedance		
12906	2023	All other reportable incidents	36241	Pacemaker/defibrillator lead	CT1684	Cardiac pacemakers and associated devices	E1906	Unspecified Infection	A24	No Apparent Device Problem	12/04/2023	12/04/2023
							F08	Hospitalization or Prolonged Hospitalization	A24	No Apparent Device Problem	12/04/2023	12/04/2023
							F1903	Device Explantation	A24	No Apparent Device Problem	12/04/2023	12/04/2023
12907	2023	All other reportable incidents	36241	Pacemaker/defibrillator lead	CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070909	Oversensing		
									A090801	Signal Artifact/Noise		
							F26	No Health Consequences or Impact	A070909	Oversensing		
									A090801	Signal Artifact/Noise		
12908	2023		18111	Pacing lead extraction kit, reusable	CT1684	Cardiac pacemakers and associated devices	F05	Delay to Treatment/ Therapy	A0101	Patient-Device Incompatibility		
12909	2023	All other reportable incidents	37265	Dual-chamber implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A070602	Delayed Charge Time		
									A070603	Failure to Charge		

							F02	Death	A070602	Delayed Charge Time		
									A070603	Failure to Charge		
12910	2023	All other reportable incidents	47265	Dual-chamber implantable pacemaker, rate-responsive	CT1685	Cardiac pacemakers	E011902	Presyncope	A071102	Failure to Interrogate		
									A071204	No Pacing		
							F12	Serious Injury/ Illness/ Impairment	A071102	Failure to Interrogate		
									A071204	No Pacing		
					CT257	Implantable cardiac pacemakers	E011902	Presyncope	A071102	Failure to Interrogate		
									A071204	No Pacing		
							F12	Serious Injury/ Illness/ Impairment	A071102	Failure to Interrogate		
									A071204	No Pacing		
					CT2866	Implantable cardiac electrical stimulators	E011902	Presyncope	A071102	Failure to Interrogate		
									A071204	No Pacing		
							F12	Serious Injury/ Illness/ Impairment	A071102	Failure to Interrogate		
									A071204	No Pacing		
12911	2023		47265	Dual-chamber implantable pacemaker, rate-responsive	CT141	Electrodes/Leads and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem		
							F1901	Additional Surgery	A24	No Apparent Device Problem		
					CT145	Cardiovascular devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem		
							F1901	Additional Surgery	A24	No Apparent Device Problem		
					CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem		
							F1901	Additional Surgery	A24	No Apparent Device Problem		
					CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem		
							F1901	Additional Surgery	A24	No Apparent Device Problem		
					CT233	Surgical	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem		
							F1901	Additional Surgery	A24	No Apparent Device Problem		
					CT2406	Implantable	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem		
							F1901	Additional Surgery	A24	No Apparent Device Problem		
					CT2407	Active implantable	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem		
							F1901	Additional Surgery	A24	No Apparent Device Problem		
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem		
							F1901	Additional Surgery	A24	No Apparent Device Problem		
					CT2727	Active implant devices/ancillaries	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem		
							F1901	Additional Surgery	A24	No Apparent Device Problem		
					CT2750	Implantable cardiac electrical stimulators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem		
							F1901	Additional Surgery	A24	No Apparent Device Problem		
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem		
							F1901	Additional Surgery	A24	No Apparent Device Problem		
					CT2985	Active	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem		
							F1901	Additional Surgery	A24	No Apparent Device Problem		
					CT3001	Implantable metal	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem		
							F1901	Additional Surgery	A24	No Apparent Device Problem		
					CT3002	Implantable synthetic polymer	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem		
							F1901	Additional Surgery	A24	No Apparent Device Problem		
					CT3006	Implantable inorganic material	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem		
							F1901	Additional Surgery	A24	No Apparent Device Problem		
					CT314	Electrically powered	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem		
							F1901	Additional Surgery	A24	No Apparent Device Problem		
					CT321	Long-term surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem		
							F1901	Additional Surgery	A24	No Apparent Device Problem		
					CT449	Generators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem		
							F1901	Additional Surgery	A24	No Apparent Device Problem		
					CT981	Single use	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem		
							F1901	Additional Surgery	A24	No Apparent Device Problem		
					CT983	Surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem		
							F1901	Additional Surgery	A24	No Apparent Device Problem		
12912	2023	All other reportable incidents	35039	Epicardial pacing lead	CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A040101	Fracture		
									A070909	Oversensing		
									A090801	Signal Artifact/Noise		
							F26	No Health Consequences or Impact	A040101	Fracture		
									A070909	Oversensing		
									A090801	Signal Artifact/Noise		
12913	2023	All other reportable incidents	47265	Dual-chamber implantable pacemaker, rate-responsive	CT141	Electrodes/Leads and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	23/02/2023	23/02/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	23/02/2023	23/02/2023
					CT145	Cardiovascular devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	23/02/2023	23/02/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	23/02/2023	23/02/2023
					CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	23/02/2023	23/02/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	23/02/2023	23/02/2023
					CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	23/02/2023	23/02/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	23/02/2023	23/02/2023
					CT233	Surgical	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	23/02/2023	23/02/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	23/02/2023	23/02/2023
					CT2406	Implantable	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	23/02/2023	23/02/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	23/02/2023	23/02/2023
					CT2407	Active implantable	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	23/02/2023	23/02/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	23/02/2023	23/02/2023
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	23/02/2023	23/02/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	23/02/2023	23/02/2023
					CT2727	Active implant devices/ancillaries	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	23/02/2023	23/02/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	23/02/2023	23/02/2023
					CT2750	Implantable cardiac electrical stimulators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	23/02/2023	23/02/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	23/02/2023	23/02/2023
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	23/02/2023	23/02/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	23/02/2023	23/02/2023
					CT2985	Active	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	23/02/2023	23/02/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	23/02/2023	23/02/2023
					CT3001	Implantable metal	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	23/02/2023	23/02/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	23/02/2023	23/02/2023
					CT3002	Implantable synthetic polymer	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	23/02/2023	23/02/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	23/02/2023	23/02/2023

					CT3006	Implantable inorganic material	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	23/02/2023	23/02/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	23/02/2023	23/02/2023
					CT314	Electrically powered	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	23/02/2023	23/02/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	23/02/2023	23/02/2023
					CT321	Long-term surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	23/02/2023	23/02/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	23/02/2023	23/02/2023
					CT449	Generators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	23/02/2023	23/02/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	23/02/2023	23/02/2023
					CT981	Single use	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	23/02/2023	23/02/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	23/02/2023	23/02/2023
					CT983	Surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	23/02/2023	23/02/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	23/02/2023	23/02/2023
12914	2023	All other reportable incidents	47265	Dual-chamber implantable pacemaker, rate-responsive	CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	14/03/2023	14/03/2023
									A071205	Pacemaker Found in Back-Up Mode	14/03/2023	14/03/2023
							F1901	Additional Surgery	A070504	Premature Discharge of Battery	14/03/2023	14/03/2023
									A071205	Pacemaker Found in Back-Up Mode	14/03/2023	14/03/2023
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	14/03/2023	14/03/2023
									A071205	Pacemaker Found in Back-Up Mode	14/03/2023	14/03/2023
							F1901	Additional Surgery	A070504	Premature Discharge of Battery	14/03/2023	14/03/2023
									A071205	Pacemaker Found in Back-Up Mode	14/03/2023	14/03/2023
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	14/03/2023	14/03/2023
									A071205	Pacemaker Found in Back-Up Mode	14/03/2023	14/03/2023
							F1901	Additional Surgery	A070504	Premature Discharge of Battery	14/03/2023	14/03/2023
									A071205	Pacemaker Found in Back-Up Mode	14/03/2023	14/03/2023
12915	2023	All other reportable incidents	60789	Intracardiac pacemaker	CT1685	Cardiac pacemakers	E020204	Malaise	A24	No Apparent Device Problem		
							E050304	Thromboembolism	A24	No Apparent Device Problem		
							E0610	Endocarditis	A24	No Apparent Device Problem		
							E2007	Fall	A24	No Apparent Device Problem		
							E2205	Bacteremia	A24	No Apparent Device Problem		
							E230101	Fever	A24	No Apparent Device Problem		
							F08	Hospitalization or Prolonged Hospitalization	A24	No Apparent Device Problem		
							F1203	Life Threatening Illness or Injury	A24	No Apparent Device Problem		
							F1903	Device Explantation	A24	No Apparent Device Problem		
							F2303	Medication Required	A24	No Apparent Device Problem		
					CT257	Implantable cardiac pacemakers	E020204	Malaise	A24	No Apparent Device Problem		
							E050304	Thromboembolism	A24	No Apparent Device Problem		
							E0610	Endocarditis	A24	No Apparent Device Problem		
							E2007	Fall	A24	No Apparent Device Problem		
							E2205	Bacteremia	A24	No Apparent Device Problem		
							E230101	Fever	A24	No Apparent Device Problem		
							F08	Hospitalization or Prolonged Hospitalization	A24	No Apparent Device Problem		
							F1203	Life Threatening Illness or Injury	A24	No Apparent Device Problem		
							F1903	Device Explantation	A24	No Apparent Device Problem		
							F2303	Medication Required	A24	No Apparent Device Problem		
					CT2818	Catheterization/Cannulation/Fluid path devices	E020204	Malaise	A24	No Apparent Device Problem		
							E050304	Thromboembolism	A24	No Apparent Device Problem		
							E0610	Endocarditis	A24	No Apparent Device Problem		
							E2007	Fall	A24	No Apparent Device Problem		
							E2205	Bacteremia	A24	No Apparent Device Problem		
							E230101	Fever	A24	No Apparent Device Problem		
							F08	Hospitalization or Prolonged Hospitalization	A24	No Apparent Device Problem		
							F1203	Life Threatening Illness or Injury	A24	No Apparent Device Problem		
							F1903	Device Explantation	A24	No Apparent Device Problem		
							F2303	Medication Required	A24	No Apparent Device Problem		
					CT2866	Implantable cardiac electrical stimulators	E020204	Malaise	A24	No Apparent Device Problem		
							E050304	Thromboembolism	A24	No Apparent Device Problem		
							E0610	Endocarditis	A24	No Apparent Device Problem		
							E2007	Fall	A24	No Apparent Device Problem		
							E2205	Bacteremia	A24	No Apparent Device Problem		
							E230101	Fever	A24	No Apparent Device Problem		
							F08	Hospitalization or Prolonged Hospitalization	A24	No Apparent Device Problem		
							F1203	Life Threatening Illness or Injury	A24	No Apparent Device Problem		
							F1903	Device Explantation	A24	No Apparent Device Problem		
							F2303	Medication Required	A24	No Apparent Device Problem		
12916	2023	All other reportable incidents	17882	Physiologic-monitoring defibrillation system	CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A0719	Unexpected Shutdown		
							F27	Problem Identified During Non-Clinical Procedure	A0719	Unexpected Shutdown		
12917	2023	All other reportable incidents	47265	Dual-chamber implantable pacemaker, rate-responsive	CT141	Electrodes/Leads and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT145	Cardiovascular devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT233	Surgical	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT2406	Implantable	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT2407	Active implantable	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT2727	Active implant devices/ancillaries	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT2750	Implantable cardiac electrical stimulators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		

					CT2985	Active	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT3001	Implantable metal	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT3002	Implantable synthetic polymer	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT3006	Implantable inorganic material	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT314	Electrically powered	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT321	Long-term surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT449	Generators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT981	Single use	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT983	Surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
12918	2023	All other reportable incidents	17882	Physiologic-monitoring defibrillation system	CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070803	Failure to Power Up		
							F27	Problem Identified During Non-Clinical Procedure	A070803	Failure to Power Up		
12919	2023	All other reportable incidents	35223	Endocardial/interventricular septal pacing lead	CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070909	Oversensing	31/03/2023	31/03/2023
							F1901	Additional Surgery	A070909	Oversensing	31/03/2023	31/03/2023
12920	2023	All other reportable incidents	17882	Physiologic-monitoring defibrillation system	CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A0719	Unexpected Shutdown		
							F27	Problem Identified During Non-Clinical Procedure	A0719	Unexpected Shutdown		
12921	2023	All other reportable incidents	35852	Single-chamber implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E2104	Electric Shock	A070909	Oversensing		
									A071302	Inappropriate/Inadequate Shock/Stimulation		
									A090801	Signal Artifact/Noise		
							F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A070909	Oversensing		
									A071302	Inappropriate/Inadequate Shock/Stimulation		
									A090801	Signal Artifact/Noise		
12922	2023		47265	Dual-chamber implantable pacemaker, rate-responsive	CT141	Electrodes/Leads and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	01/03/2023	01/03/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	01/03/2023	01/03/2023
					CT145	Cardiovascular devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	01/03/2023	01/03/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	01/03/2023	01/03/2023
					CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	01/03/2023	01/03/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	01/03/2023	01/03/2023
					CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	01/03/2023	01/03/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	01/03/2023	01/03/2023
					CT233	Surgical	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	01/03/2023	01/03/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	01/03/2023	01/03/2023
					CT2406	Implantable	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	01/03/2023	01/03/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	01/03/2023	01/03/2023
					CT2407	Active implantable	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	01/03/2023	01/03/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	01/03/2023	01/03/2023
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	01/03/2023	01/03/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	01/03/2023	01/03/2023
					CT2727	Active implant devices/ancillaries	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	01/03/2023	01/03/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	01/03/2023	01/03/2023
					CT2750	Implantable cardiac electrical stimulators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	01/03/2023	01/03/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	01/03/2023	01/03/2023
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	01/03/2023	01/03/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	01/03/2023	01/03/2023
					CT2985	Active	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	01/03/2023	01/03/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	01/03/2023	01/03/2023
					CT3001	Implantable metal	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	01/03/2023	01/03/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	01/03/2023	01/03/2023
					CT3002	Implantable synthetic polymer	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	01/03/2023	01/03/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	01/03/2023	01/03/2023
					CT3006	Implantable inorganic material	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	01/03/2023	01/03/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	01/03/2023	01/03/2023
					CT314	Electrically powered	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	01/03/2023	01/03/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	01/03/2023	01/03/2023
					CT321	Long-term surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	01/03/2023	01/03/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	01/03/2023	01/03/2023
					CT449	Generators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	01/03/2023	01/03/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	01/03/2023	01/03/2023
					CT981	Single use	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	01/03/2023	01/03/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	01/03/2023	01/03/2023
					CT983	Surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	01/03/2023	01/03/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	01/03/2023	01/03/2023
12923	2023	All other reportable incidents	35852	Single-chamber implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	25/04/2023	25/04/2023
							F1901	Additional Surgery	A070504	Premature Discharge of Battery	25/04/2023	25/04/2023
12924	2023	All other reportable incidents	17882	Physiologic-monitoring defibrillation system	CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A07	Electrical /Electronic Property Problem		
							F27	Problem Identified During Non-Clinical Procedure	A07	Electrical /Electronic Property Problem		
12925	2023	All other reportable incidents	17882	Physiologic-monitoring defibrillation system	CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A0908	Incorrect, Inadequate or Imprecise Result or Readings		
							F27	Problem Identified During Non-Clinical Procedure	A0908	Incorrect, Inadequate or Imprecise Result or Readings		
12926	2023	All other reportable incidents	17882	Physiologic-monitoring defibrillation system	CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A0908	Incorrect, Inadequate or Imprecise Result or Readings		
							F27	Problem Identified During Non-Clinical Procedure	A0908	Incorrect, Inadequate or Imprecise Result or Readings		
12927	2023		47265	Dual-chamber implantable pacemaker, rate-responsive	CT141	Electrodes/Leads and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	17/02/2023	17/02/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	17/02/2023	17/02/2023
					CT145	Cardiovascular devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	17/02/2023	17/02/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	17/02/2023	17/02/2023
					CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	17/02/2023	17/02/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	17/02/2023	17/02/2023
					CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	17/02/2023	17/02/2023

							F1901	Additional Surgery	A24	No Apparent Device Problem	17/02/2023	17/02/2023
					CT233	Surgical	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	17/02/2023	17/02/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	17/02/2023	17/02/2023
					CT2406	Implantable	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	17/02/2023	17/02/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	17/02/2023	17/02/2023
					CT2407	Active implantable	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	17/02/2023	17/02/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	17/02/2023	17/02/2023
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	17/02/2023	17/02/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	17/02/2023	17/02/2023
					CT2727	Active implant devices/ancillaries	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	17/02/2023	17/02/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	17/02/2023	17/02/2023
					CT2750	Implantable cardiac electrical stimulators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	17/02/2023	17/02/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	17/02/2023	17/02/2023
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	17/02/2023	17/02/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	17/02/2023	17/02/2023
					CT2985	Active	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	17/02/2023	17/02/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	17/02/2023	17/02/2023
					CT3001	Implantable metal	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	17/02/2023	17/02/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	17/02/2023	17/02/2023
					CT3002	Implantable synthetic polymer	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	17/02/2023	17/02/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	17/02/2023	17/02/2023
					CT3006	Implantable inorganic material	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	17/02/2023	17/02/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	17/02/2023	17/02/2023
					CT314	Electrically powered	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	17/02/2023	17/02/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	17/02/2023	17/02/2023
					CT321	Long-term surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	17/02/2023	17/02/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	17/02/2023	17/02/2023
					CT449	Generators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	17/02/2023	17/02/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	17/02/2023	17/02/2023
					CT981	Single use	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	17/02/2023	17/02/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	17/02/2023	17/02/2023
					CT983	Surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	17/02/2023	17/02/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	17/02/2023	17/02/2023
12928	2023	All other reportable incidents	47263	Cardiac resynchronization therapy implantable pacemaker	CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A0908	Incorrect, Inadequate or Imprecise Result or Readings		
							F26	No Health Consequences or Impact	A0908	Incorrect, Inadequate or Imprecise Result or Readings		
					CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A0908	Incorrect, Inadequate or Imprecise Result or Readings		
							F26	No Health Consequences or Impact	A0908	Incorrect, Inadequate or Imprecise Result or Readings		
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A0908	Incorrect, Inadequate or Imprecise Result or Readings		
							F26	No Health Consequences or Impact	A0908	Incorrect, Inadequate or Imprecise Result or Readings		
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A0908	Incorrect, Inadequate or Imprecise Result or Readings		
							F26	No Health Consequences or Impact	A0908	Incorrect, Inadequate or Imprecise Result or Readings		
12929	2023		47265	Dual-chamber implantable pacemaker, rate-responsive	CT141	Electrodes/Leads and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	10/02/2023	10/02/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	10/02/2023	10/02/2023
					CT145	Cardiovascular devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	10/02/2023	10/02/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	10/02/2023	10/02/2023
					CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	10/02/2023	10/02/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	10/02/2023	10/02/2023
					CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	10/02/2023	10/02/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	10/02/2023	10/02/2023
					CT233	Surgical	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	10/02/2023	10/02/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	10/02/2023	10/02/2023
					CT2406	Implantable	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	10/02/2023	10/02/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	10/02/2023	10/02/2023
					CT2407	Active implantable	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	10/02/2023	10/02/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	10/02/2023	10/02/2023
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	10/02/2023	10/02/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	10/02/2023	10/02/2023
					CT2727	Active implant devices/ancillaries	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	10/02/2023	10/02/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	10/02/2023	10/02/2023
					CT2750	Implantable cardiac electrical stimulators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	10/02/2023	10/02/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	10/02/2023	10/02/2023
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	10/02/2023	10/02/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	10/02/2023	10/02/2023
					CT2985	Active	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	10/02/2023	10/02/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	10/02/2023	10/02/2023
					CT3001	Implantable metal	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	10/02/2023	10/02/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	10/02/2023	10/02/2023
					CT3002	Implantable synthetic polymer	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	10/02/2023	10/02/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	10/02/2023	10/02/2023
					CT3006	Implantable inorganic material	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	10/02/2023	10/02/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	10/02/2023	10/02/2023
					CT314	Electrically powered	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	10/02/2023	10/02/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	10/02/2023	10/02/2023
					CT321	Long-term surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	10/02/2023	10/02/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	10/02/2023	10/02/2023
					CT449	Generators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	10/02/2023	10/02/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	10/02/2023	10/02/2023
					CT981	Single use	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	10/02/2023	10/02/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	10/02/2023	10/02/2023
					CT983	Surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	10/02/2023	10/02/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	10/02/2023	10/02/2023
12930	2023	All other reportable incidents	47265	Dual-chamber implantable pacemaker, rate-	CT141	Electrodes/Leads and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	19/03/2023	19/03/2023

				responsive			F1901	Additional Surgery	A24	No Apparent Device Problem	19/03/2023	19/03/2023
					CT145	Cardiovascular devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	19/03/2023	19/03/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	19/03/2023	19/03/2023
					CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	19/03/2023	19/03/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	19/03/2023	19/03/2023
					CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	19/03/2023	19/03/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	19/03/2023	19/03/2023
					CT233	Surgical	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	19/03/2023	19/03/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	19/03/2023	19/03/2023
					CT2406	Implantable	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	19/03/2023	19/03/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	19/03/2023	19/03/2023
					CT2407	Active implantable	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	19/03/2023	19/03/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	19/03/2023	19/03/2023
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	19/03/2023	19/03/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	19/03/2023	19/03/2023
					CT2727	Active implant devices/ancillaries	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	19/03/2023	19/03/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	19/03/2023	19/03/2023
					CT2750	Implantable cardiac electrical stimulators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	19/03/2023	19/03/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	19/03/2023	19/03/2023
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	19/03/2023	19/03/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	19/03/2023	19/03/2023
					CT2985	Active	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	19/03/2023	19/03/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	19/03/2023	19/03/2023
					CT3001	Implantable metal	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	19/03/2023	19/03/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	19/03/2023	19/03/2023
					CT3002	Implantable synthetic polymer	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	19/03/2023	19/03/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	19/03/2023	19/03/2023
					CT3006	Implantable inorganic material	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	19/03/2023	19/03/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	19/03/2023	19/03/2023
					CT314	Electrically powered	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	19/03/2023	19/03/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	19/03/2023	19/03/2023
					CT321	Long-term surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	19/03/2023	19/03/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	19/03/2023	19/03/2023
					CT449	Generators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	19/03/2023	19/03/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	19/03/2023	19/03/2023
					CT981	Single use	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	19/03/2023	19/03/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	19/03/2023	19/03/2023
					CT983	Surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	19/03/2023	19/03/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	19/03/2023	19/03/2023
12931	2023		47265	Dual-chamber implantable pacemaker, rate-responsive	CT141	Electrodes/Leads and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	12/04/2023	12/04/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	12/04/2023	12/04/2023
					CT145	Cardiovascular devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	12/04/2023	12/04/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	12/04/2023	12/04/2023
					CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	12/04/2023	12/04/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	12/04/2023	12/04/2023
					CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	12/04/2023	12/04/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	12/04/2023	12/04/2023
					CT233	Surgical	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	12/04/2023	12/04/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	12/04/2023	12/04/2023
					CT2406	Implantable	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	12/04/2023	12/04/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	12/04/2023	12/04/2023
					CT2407	Active implantable	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	12/04/2023	12/04/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	12/04/2023	12/04/2023
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	12/04/2023	12/04/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	12/04/2023	12/04/2023
					CT2727	Active implant devices/ancillaries	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	12/04/2023	12/04/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	12/04/2023	12/04/2023
					CT2750	Implantable cardiac electrical stimulators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	12/04/2023	12/04/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	12/04/2023	12/04/2023
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	12/04/2023	12/04/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	12/04/2023	12/04/2023
					CT2985	Active	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	12/04/2023	12/04/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	12/04/2023	12/04/2023
					CT3001	Implantable metal	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	12/04/2023	12/04/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	12/04/2023	12/04/2023
					CT3002	Implantable synthetic polymer	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	12/04/2023	12/04/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	12/04/2023	12/04/2023
					CT3006	Implantable inorganic material	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	12/04/2023	12/04/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	12/04/2023	12/04/2023
					CT314	Electrically powered	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	12/04/2023	12/04/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	12/04/2023	12/04/2023
					CT321	Long-term surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	12/04/2023	12/04/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	12/04/2023	12/04/2023
					CT449	Generators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	12/04/2023	12/04/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	12/04/2023	12/04/2023
					CT981	Single use	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	12/04/2023	12/04/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	12/04/2023	12/04/2023
					CT983	Surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	12/04/2023	12/04/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	12/04/2023	12/04/2023
12932	2023		47265	Dual-chamber implantable pacemaker, rate-responsive	CT141	Electrodes/Leads and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem		
							F1901	Additional Surgery	A24	No Apparent Device Problem		
					CT145	Cardiovascular devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem		
							F1901	Additional Surgery	A24	No Apparent Device Problem		
					CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem		
							F1901	Additional Surgery	A24	No Apparent Device Problem		
					CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem		
							F1901	Additional Surgery	A24	No Apparent Device Problem		
					CT233	Surgical	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem		

							F1901	Additional Surgery	A24	No Apparent Device Problem		
					CT2406	Implantable	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem		
							F1901	Additional Surgery	A24	No Apparent Device Problem		
					CT2407	Active implantable	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem		
							F1901	Additional Surgery	A24	No Apparent Device Problem		
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem		
							F1901	Additional Surgery	A24	No Apparent Device Problem		
					CT2727	Active implant devices/ancillaries	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem		
							F1901	Additional Surgery	A24	No Apparent Device Problem		
					CT2750	Implantable cardiac electrical stimulators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem		
							F1901	Additional Surgery	A24	No Apparent Device Problem		
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem		
							F1901	Additional Surgery	A24	No Apparent Device Problem		
					CT2985	Active	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem		
							F1901	Additional Surgery	A24	No Apparent Device Problem		
					CT3001	Implantable metal	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem		
							F1901	Additional Surgery	A24	No Apparent Device Problem		
					CT3002	Implantable synthetic polymer	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem		
							F1901	Additional Surgery	A24	No Apparent Device Problem		
					CT3006	Implantable inorganic material	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem		
							F1901	Additional Surgery	A24	No Apparent Device Problem		
					CT314	Electrically powered	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem		
							F1901	Additional Surgery	A24	No Apparent Device Problem		
					CT321	Long-term surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem		
							F1901	Additional Surgery	A24	No Apparent Device Problem		
					CT449	Generators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem		
							F1901	Additional Surgery	A24	No Apparent Device Problem		
					CT981	Single use	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem		
							F1901	Additional Surgery	A24	No Apparent Device Problem		
					CT983	Surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem		
							F1901	Additional Surgery	A24	No Apparent Device Problem		
12933	2023	All other reportable incidents	37265	Dual-chamber implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E2104	Electric Shock	A070909	Oversensing		
									A071302	Inappropriate/Inadequate Shock/Stimulation		
									A090801	Signal Artifact/Noise		
									A1301	Failure to Read Input Signal		
							F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A070909	Oversensing		
									A071302	Inappropriate/Inadequate Shock/Stimulation		
									A090801	Signal Artifact/Noise		
									A1301	Failure to Read Input Signal		
12934	2023	All other reportable incidents	47265	Dual-chamber implantable pacemaker, rate-responsive	CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
12935	2023		47265	Dual-chamber implantable pacemaker, rate-responsive	CT141	Electrodes/Leads and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem		
							F1901	Additional Surgery	A24	No Apparent Device Problem		
					CT145	Cardiovascular devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem		
							F1901	Additional Surgery	A24	No Apparent Device Problem		
					CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem		
							F1901	Additional Surgery	A24	No Apparent Device Problem		
					CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem		
							F1901	Additional Surgery	A24	No Apparent Device Problem		
					CT233	Surgical	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem		
							F1901	Additional Surgery	A24	No Apparent Device Problem		
					CT2406	Implantable	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem		
							F1901	Additional Surgery	A24	No Apparent Device Problem		
					CT2407	Active implantable	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem		
							F1901	Additional Surgery	A24	No Apparent Device Problem		
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem		
							F1901	Additional Surgery	A24	No Apparent Device Problem		
					CT2727	Active implant devices/ancillaries	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem		
							F1901	Additional Surgery	A24	No Apparent Device Problem		
					CT2750	Implantable cardiac electrical stimulators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem		
							F1901	Additional Surgery	A24	No Apparent Device Problem		
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem		
							F1901	Additional Surgery	A24	No Apparent Device Problem		
					CT2985	Active	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem		
							F1901	Additional Surgery	A24	No Apparent Device Problem		
					CT3001	Implantable metal	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem		
							F1901	Additional Surgery	A24	No Apparent Device Problem		
					CT3002	Implantable synthetic polymer	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem		
							F1901	Additional Surgery	A24	No Apparent Device Problem		
					CT3006	Implantable inorganic material	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem		
							F1901	Additional Surgery	A24	No Apparent Device Problem		
					CT314	Electrically powered	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem		
							F1901	Additional Surgery	A24	No Apparent Device Problem		
					CT321	Long-term surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem		
							F1901	Additional Surgery	A24	No Apparent Device Problem		
					CT449	Generators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem		
							F1901	Additional Surgery	A24	No Apparent Device Problem		
					CT981	Single use	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem		

							F1901	Additional Surgery	A24	No Apparent Device Problem		
					CT983	Surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem		
							F1901	Additional Surgery	A24	No Apparent Device Problem		
12936	2023	All other reportable incidents	47263	Cardiac resynchronization therapy implantable pacemaker	CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A0908	Incorrect, Inadequate or Imprecise Result or Readings	26/05/2023	26/05/2023
							F08	Hospitalization or Prolonged Hospitalization	A0908	Incorrect, Inadequate or Imprecise Result or Readings	26/05/2023	26/05/2023
							F1905	Device Revision or Replacement	A0908	Incorrect, Inadequate or Imprecise Result or Readings	26/05/2023	26/05/2023
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A0908	Incorrect, Inadequate or Imprecise Result or Readings	26/05/2023	26/05/2023
							F08	Hospitalization or Prolonged Hospitalization	A0908	Incorrect, Inadequate or Imprecise Result or Readings	26/05/2023	26/05/2023
							F1905	Device Revision or Replacement	A0908	Incorrect, Inadequate or Imprecise Result or Readings	26/05/2023	26/05/2023
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A0908	Incorrect, Inadequate or Imprecise Result or Readings	26/05/2023	26/05/2023
							F08	Hospitalization or Prolonged Hospitalization	A0908	Incorrect, Inadequate or Imprecise Result or Readings	26/05/2023	26/05/2023
							F1905	Device Revision or Replacement	A0908	Incorrect, Inadequate or Imprecise Result or Readings	26/05/2023	26/05/2023
12937	2023	All other reportable incidents	35852	Single-chamber implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A0508	Noise, Audible		
									A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A0508	Noise, Audible		
									A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
12938	2023		47265	Dual-chamber implantable pacemaker, rate-responsive			E2403	No Clinical Signs, Symptoms or Conditions	A090801	Signal Artifact/Noise	21/04/2023	21/04/2023
									A1305	Wireless Communication Problem	21/04/2023	21/04/2023
							F1901	Additional Surgery	A090801	Signal Artifact/Noise	21/04/2023	21/04/2023
									A1305	Wireless Communication Problem	21/04/2023	21/04/2023
12939	2023	All other reportable incidents	47265	Dual-chamber implantable pacemaker, rate-responsive	CT141	Electrodes/Leads and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A072201	High Impedance		
							F26	No Health Consequences or Impact	A072201	High Impedance		
					CT145	Cardiovascular devices	E2403	No Clinical Signs, Symptoms or Conditions	A072201	High Impedance		
							F26	No Health Consequences or Impact	A072201	High Impedance		
					CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A072201	High Impedance		
							F26	No Health Consequences or Impact	A072201	High Impedance		
					CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A072201	High Impedance		
							F26	No Health Consequences or Impact	A072201	High Impedance		
					CT233	Surgical	E2403	No Clinical Signs, Symptoms or Conditions	A072201	High Impedance		
							F26	No Health Consequences or Impact	A072201	High Impedance		
					CT2406	Implantable	E2403	No Clinical Signs, Symptoms or Conditions	A072201	High Impedance		
							F26	No Health Consequences or Impact	A072201	High Impedance		
					CT2407	Active implantable	E2403	No Clinical Signs, Symptoms or Conditions	A072201	High Impedance		
							F26	No Health Consequences or Impact	A072201	High Impedance		
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A072201	High Impedance		
							F26	No Health Consequences or Impact	A072201	High Impedance		
					CT2727	Active implant devices/ancillaries	E2403	No Clinical Signs, Symptoms or Conditions	A072201	High Impedance		
							F26	No Health Consequences or Impact	A072201	High Impedance		
					CT2750	Implantable cardiac electrical stimulators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A072201	High Impedance		
							F26	No Health Consequences or Impact	A072201	High Impedance		
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A072201	High Impedance		
							F26	No Health Consequences or Impact	A072201	High Impedance		
					CT2985	Active	E2403	No Clinical Signs, Symptoms or Conditions	A072201	High Impedance		
							F26	No Health Consequences or Impact	A072201	High Impedance		
					CT3001	Implantable metal	E2403	No Clinical Signs, Symptoms or Conditions	A072201	High Impedance		
							F26	No Health Consequences or Impact	A072201	High Impedance		
					CT3002	Implantable synthetic polymer	E2403	No Clinical Signs, Symptoms or Conditions	A072201	High Impedance		
							F26	No Health Consequences or Impact	A072201	High Impedance		
					CT3006	Implantable inorganic material	E2403	No Clinical Signs, Symptoms or Conditions	A072201	High Impedance		
							F26	No Health Consequences or Impact	A072201	High Impedance		
					CT314	Electrically powered	E2403	No Clinical Signs, Symptoms or Conditions	A072201	High Impedance		
							F26	No Health Consequences or Impact	A072201	High Impedance		
					CT321	Long-term surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A072201	High Impedance		
							F26	No Health Consequences or Impact	A072201	High Impedance		
					CT449	Generators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A072201	High Impedance		
							F26	No Health Consequences or Impact	A072201	High Impedance		
					CT981	Single use	E2403	No Clinical Signs, Symptoms or Conditions	A072201	High Impedance		
							F26	No Health Consequences or Impact	A072201	High Impedance		
					CT983	Surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A072201	High Impedance		
							F26	No Health Consequences or Impact	A072201	High Impedance		
12940	2023	All other reportable incidents	35852	Single-chamber implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
									A0908	Incorrect, Inadequate or Imprecise Result or Readings		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
									A0908	Incorrect, Inadequate or Imprecise Result or Readings		
12941	2023	All other reportable incidents	47267	Single-chamber implantable pacemaker, rate-responsive	CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	21/04/2023	21/04/2023
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	21/04/2023	21/04/2023
							F12	Serious Injury/ Illness/ Impairment	A070504	Premature Discharge of Battery	21/04/2023	21/04/2023
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	21/04/2023	21/04/2023
							F23	Unexpected Medical Intervention	A070504	Premature Discharge of Battery	21/04/2023	21/04/2023
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	21/04/2023	21/04/2023
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	21/04/2023	21/04/2023
							F12	Serious Injury/ Illness/ Impairment	A070504	Premature Discharge of Battery	21/04/2023	21/04/2023
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	21/04/2023	21/04/2023
							F23	Unexpected Medical Intervention	A070504	Premature Discharge of Battery	21/04/2023	21/04/2023
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	21/04/2023	21/04/2023
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	21/04/2023	21/04/2023
							F12	Serious Injury/ Illness/ Impairment	A070504	Premature Discharge of Battery	21/04/2023	21/04/2023

							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	21/04/2023	21/04/2023
							F23	Unexpected Medical Intervention	A070504	Premature Discharge of Battery	21/04/2023	21/04/2023
12942	2023	All other reportable incidents	35223	Endocardial/interventricular septal pacing lead	CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A0709	Device Sensing Problem		
									A070901	Decreased Sensitivity		
									A070910	Undersensing		
							F26	No Health Consequences or Impact	A0709	Device Sensing Problem		
									A070901	Decreased Sensitivity		
									A070910	Undersensing		
12943	2023	All other reportable incidents	35852	Single-chamber implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E2104	Electric Shock	A070909	Oversensing		
									A071302	Inappropriate/Inadequate Shock/Stimulation		
									A090801	Signal Artifact/Noise		
							F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A070909	Oversensing		
									A071302	Inappropriate/Inadequate Shock/Stimulation		
									A090801	Signal Artifact/Noise		
12944	2023	All other reportable incidents	47270	Cardiac resynchronization therapy implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
12945	2023	All other reportable incidents	47265	Dual-chamber implantable pacemaker, rate-responsive	CT141	Electrodes/Leads and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT145	Cardiovascular devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT233	Surgical	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT2406	Implantable	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT2407	Active implantable	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT2727	Active implant devices/ancillaries	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT2750	Implantable cardiac electrical stimulators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT2985	Active	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT3001	Implantable metal	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT3002	Implantable synthetic polymer	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT3006	Implantable inorganic material	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT314	Electrically powered	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT321	Long-term surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT449	Generators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT981	Single use	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT983	Surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
12946	2023	All other reportable incidents	37265	Dual-chamber implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E2104	Electric Shock	A070909	Oversensing	20/04/2023	20/04/2023
									A071302	Inappropriate/Inadequate Shock/Stimulation	20/04/2023	20/04/2023
							F08	Hospitalization or Prolonged Hospitalization	A070909	Oversensing	20/04/2023	20/04/2023
									A071302	Inappropriate/Inadequate Shock/Stimulation	20/04/2023	20/04/2023
							F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A070909	Oversensing	20/04/2023	20/04/2023
									A071302	Inappropriate/Inadequate Shock/Stimulation	20/04/2023	20/04/2023
							F1905	Device Revision or Replacement	A070909	Oversensing	20/04/2023	20/04/2023
									A071302	Inappropriate/Inadequate Shock/Stimulation	20/04/2023	20/04/2023
							F23	Unexpected Medical Intervention	A070909	Oversensing	20/04/2023	20/04/2023
									A071302	Inappropriate/Inadequate Shock/Stimulation	20/04/2023	20/04/2023
12947	2023	All other reportable incidents	17882	Physiologic-monitoring defibrillation system	CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A0902	Display or Visual Feedback Problem		
							F27	Problem Identified During Non-Clinical Procedure	A0902	Display or Visual Feedback Problem		
12948	2023	All other reportable incidents	47265	Dual-chamber implantable pacemaker, rate-responsive	CT141	Electrodes/Leads and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT145	Cardiovascular devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT233	Surgical	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT2406	Implantable	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT2407	Active implantable	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT2727	Active implant devices/ancillaries	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		

					CT2750	Implantable cardiac electrical stimulators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT2985	Active	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT3001	Implantable metal	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT3002	Implantable synthetic polymer	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT3006	Implantable inorganic material	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT314	Electrically powered	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT321	Long-term surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT449	Generators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT981	Single use	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT983	Surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
12949	2023	All other reportable incidents	47263	Cardiac resynchronization therapy implantable pacemaker	CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A0908	Incorrect, Inadequate or Imprecise Result or Readings	23/12/2022	23/12/2022
							F08	Hospitalization or Prolonged Hospitalization	A0908	Incorrect, Inadequate or Imprecise Result or Readings	23/12/2022	23/12/2022
							F1905	Device Revision or Replacement	A0908	Incorrect, Inadequate or Imprecise Result or Readings	23/12/2022	23/12/2022
					CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A0908	Incorrect, Inadequate or Imprecise Result or Readings	23/12/2022	23/12/2022
							F08	Hospitalization or Prolonged Hospitalization	A0908	Incorrect, Inadequate or Imprecise Result or Readings	23/12/2022	23/12/2022
							F1905	Device Revision or Replacement	A0908	Incorrect, Inadequate or Imprecise Result or Readings	23/12/2022	23/12/2022
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A0908	Incorrect, Inadequate or Imprecise Result or Readings	23/12/2022	23/12/2022
							F08	Hospitalization or Prolonged Hospitalization	A0908	Incorrect, Inadequate or Imprecise Result or Readings	23/12/2022	23/12/2022
							F1905	Device Revision or Replacement	A0908	Incorrect, Inadequate or Imprecise Result or Readings	23/12/2022	23/12/2022
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A0908	Incorrect, Inadequate or Imprecise Result or Readings	23/12/2022	23/12/2022
							F08	Hospitalization or Prolonged Hospitalization	A0908	Incorrect, Inadequate or Imprecise Result or Readings	23/12/2022	23/12/2022
							F1905	Device Revision or Replacement	A0908	Incorrect, Inadequate or Imprecise Result or Readings	23/12/2022	23/12/2022
12950	2023	All other reportable incidents	37265	Dual-chamber implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E0601	Arrhythmia	A071201	Failure to Convert Rhythm		
							F12	Serious Injury/ Illness/ Impairment	A071201	Failure to Convert Rhythm		
12951	2023	All other reportable incidents	17882	Physiologic-monitoring defibrillation system	CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A0719	Unexpected Shutdown		
							F27	Problem Identified During Non-Clinical Procedure	A0719	Unexpected Shutdown		
12952	2023		47265	Dual-chamber implantable pacemaker, rate-responsive	CT141	Electrodes/Leads and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	02/03/2023	02/03/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	02/03/2023	02/03/2023
					CT145	Cardiovascular devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	02/03/2023	02/03/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	02/03/2023	02/03/2023
					CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	02/03/2023	02/03/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	02/03/2023	02/03/2023
					CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	02/03/2023	02/03/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	02/03/2023	02/03/2023
					CT233	Surgical	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	02/03/2023	02/03/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	02/03/2023	02/03/2023
					CT2406	Implantable	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	02/03/2023	02/03/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	02/03/2023	02/03/2023
					CT2407	Active implantable	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	02/03/2023	02/03/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	02/03/2023	02/03/2023
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	02/03/2023	02/03/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	02/03/2023	02/03/2023
					CT2727	Active implant devices/ancillaries	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	02/03/2023	02/03/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	02/03/2023	02/03/2023
					CT2750	Implantable cardiac electrical stimulators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	02/03/2023	02/03/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	02/03/2023	02/03/2023
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	02/03/2023	02/03/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	02/03/2023	02/03/2023
					CT2985	Active	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	02/03/2023	02/03/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	02/03/2023	02/03/2023
					CT3001	Implantable metal	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	02/03/2023	02/03/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	02/03/2023	02/03/2023
					CT3002	Implantable synthetic polymer	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	02/03/2023	02/03/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	02/03/2023	02/03/2023
					CT3006	Implantable inorganic material	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	02/03/2023	02/03/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	02/03/2023	02/03/2023
					CT314	Electrically powered	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	02/03/2023	02/03/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	02/03/2023	02/03/2023
					CT321	Long-term surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	02/03/2023	02/03/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	02/03/2023	02/03/2023
					CT449	Generators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	02/03/2023	02/03/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	02/03/2023	02/03/2023
					CT981	Single use	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	02/03/2023	02/03/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	02/03/2023	02/03/2023
					CT983	Surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	02/03/2023	02/03/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	02/03/2023	02/03/2023
12953	2023	All other reportable incidents	17882	Physiologic-monitoring defibrillation system	CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070803	Failure to Power Up		

							F27	Problem Identified During Non-Clinical Procedure	A070803	Failure to Power Up		
12954	2023	All other reportable incidents	17882	Physiologic-monitoring defibrillation system	CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A150105	Key or Button Unresponsive/not Working		
							F27	Problem Identified During Non-Clinical Procedure	A150105	Key or Button Unresponsive/not Working		
12955	2023	All other reportable incidents	17882	Physiologic-monitoring defibrillation system	CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070803	Failure to Power Up		
							F27	Problem Identified During Non-Clinical Procedure	A070803	Failure to Power Up		
12956	2023	All other reportable incidents	47265	Dual-chamber implantable pacemaker, rate-responsive	CT141	Electrodes/Leads and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	13/04/2023	13/04/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	13/04/2023	13/04/2023
					CT145	Cardiovascular devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	13/04/2023	13/04/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	13/04/2023	13/04/2023
					CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	13/04/2023	13/04/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	13/04/2023	13/04/2023
					CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	13/04/2023	13/04/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	13/04/2023	13/04/2023
					CT233	Surgical	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	13/04/2023	13/04/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	13/04/2023	13/04/2023
					CT2406	Implantable	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	13/04/2023	13/04/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	13/04/2023	13/04/2023
					CT2407	Active implantable	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	13/04/2023	13/04/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	13/04/2023	13/04/2023
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	13/04/2023	13/04/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	13/04/2023	13/04/2023
					CT2727	Active implant devices/ancillaries	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	13/04/2023	13/04/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	13/04/2023	13/04/2023
					CT2750	Implantable cardiac electrical stimulators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	13/04/2023	13/04/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	13/04/2023	13/04/2023
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	13/04/2023	13/04/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	13/04/2023	13/04/2023
					CT2985	Active	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	13/04/2023	13/04/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	13/04/2023	13/04/2023
					CT3001	Implantable metal	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	13/04/2023	13/04/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	13/04/2023	13/04/2023
					CT3002	Implantable synthetic polymer	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	13/04/2023	13/04/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	13/04/2023	13/04/2023
					CT3006	Implantable inorganic material	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	13/04/2023	13/04/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	13/04/2023	13/04/2023
					CT314	Electrically powered	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	13/04/2023	13/04/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	13/04/2023	13/04/2023
					CT321	Long-term surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	13/04/2023	13/04/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	13/04/2023	13/04/2023
					CT449	Generators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	13/04/2023	13/04/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	13/04/2023	13/04/2023
					CT981	Single use	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	13/04/2023	13/04/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	13/04/2023	13/04/2023
					CT983	Surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	13/04/2023	13/04/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	13/04/2023	13/04/2023
12957	2023	All other reportable incidents	35223	Endocardial/interventricular septal pacing lead	CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A051201	Device Dislodged or Dislocated	25/04/2023	25/04/2023
									A070101	Failure to Capture	25/04/2023	25/04/2023
									A070901	Decreased Sensitivity	25/04/2023	25/04/2023
									A070909	Oversensing	25/04/2023	25/04/2023
									A0722	Impedance Problem	25/04/2023	25/04/2023
							F08	Hospitalization or Prolonged Hospitalization	A051201	Device Dislodged or Dislocated	25/04/2023	25/04/2023
									A070101	Failure to Capture	25/04/2023	25/04/2023
									A070901	Decreased Sensitivity	25/04/2023	25/04/2023
									A070909	Oversensing	25/04/2023	25/04/2023
									A0722	Impedance Problem	25/04/2023	25/04/2023
							F12	Serious Injury/ Illness/ Impairment	A051201	Device Dislodged or Dislocated	25/04/2023	25/04/2023
									A070101	Failure to Capture	25/04/2023	25/04/2023
									A070901	Decreased Sensitivity	25/04/2023	25/04/2023
									A070909	Oversensing	25/04/2023	25/04/2023
									A0722	Impedance Problem	25/04/2023	25/04/2023
							F1904	Device Repositioning	A051201	Device Dislodged or Dislocated	25/04/2023	25/04/2023
									A070101	Failure to Capture	25/04/2023	25/04/2023
									A070901	Decreased Sensitivity	25/04/2023	25/04/2023
									A070909	Oversensing	25/04/2023	25/04/2023
									A0722	Impedance Problem	25/04/2023	25/04/2023
							F1905	Device Revision or Replacement	A051201	Device Dislodged or Dislocated	25/04/2023	25/04/2023
									A070101	Failure to Capture	25/04/2023	25/04/2023
									A070901	Decreased Sensitivity	25/04/2023	25/04/2023
									A070909	Oversensing	25/04/2023	25/04/2023
									A0722	Impedance Problem	25/04/2023	25/04/2023
12958	2023	All other reportable incidents	37265	Dual-chamber implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E2104	Electric Shock	A0712	Pacing Problem		
									A071302	Inappropriate/Inadequate Shock/Stimulation		
							F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A0712	Pacing Problem		
									A071302	Inappropriate/Inadequate Shock/Stimulation		
12959	2023	All other reportable incidents	17882	Physiologic-monitoring defibrillation system	CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A0719	Unexpected Shutdown		
							F27	Problem Identified During Non-Clinical Procedure	A0719	Unexpected Shutdown		
12960	2023	All other reportable incidents	35852	Single-chamber implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A070909	Oversensing		
									A0712	Pacing Problem		
									A072201	High Impedance		
									A1301	Failure to Read Input Signal		
							F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A070909	Oversensing		
									A0712	Pacing Problem		
									A072201	High Impedance		
									A1301	Failure to Read Input Signal		
12961	2023	All other reportable incidents	47265	Dual-chamber implantable pacemaker, rate-responsive	CT141	Electrodes/Leads and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	12/04/2023	12/04/2023
									A0908	Incorrect, Inadequate or Imprecise Result or Readings	12/04/2023	12/04/2023
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	12/04/2023	12/04/2023

								A0908	Incorrect, Inadequate or Imprecise Result or Readings	12/04/2023	12/04/2023
						F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	12/04/2023	12/04/2023
								A0908	Incorrect, Inadequate or Imprecise Result or Readings	12/04/2023	12/04/2023
					CT145	Cardiovascular devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	12/04/2023	12/04/2023
									A0908	Incorrect, Inadequate or Imprecise Result or Readings	12/04/2023	12/04/2023
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	12/04/2023	12/04/2023
									A0908	Incorrect, Inadequate or Imprecise Result or Readings	12/04/2023	12/04/2023
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	12/04/2023	12/04/2023
									A0908	Incorrect, Inadequate or Imprecise Result or Readings	12/04/2023	12/04/2023
					CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	12/04/2023	12/04/2023
									A0908	Incorrect, Inadequate or Imprecise Result or Readings	12/04/2023	12/04/2023
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	12/04/2023	12/04/2023
									A0908	Incorrect, Inadequate or Imprecise Result or Readings	12/04/2023	12/04/2023
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	12/04/2023	12/04/2023
									A0908	Incorrect, Inadequate or Imprecise Result or Readings	12/04/2023	12/04/2023
					CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	12/04/2023	12/04/2023
									A0908	Incorrect, Inadequate or Imprecise Result or Readings	12/04/2023	12/04/2023
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	12/04/2023	12/04/2023
									A0908	Incorrect, Inadequate or Imprecise Result or Readings	12/04/2023	12/04/2023
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	12/04/2023	12/04/2023
									A0908	Incorrect, Inadequate or Imprecise Result or Readings	12/04/2023	12/04/2023
					CT233	Surgical	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	12/04/2023	12/04/2023
									A0908	Incorrect, Inadequate or Imprecise Result or Readings	12/04/2023	12/04/2023
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	12/04/2023	12/04/2023
									A0908	Incorrect, Inadequate or Imprecise Result or Readings	12/04/2023	12/04/2023
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	12/04/2023	12/04/2023
									A0908	Incorrect, Inadequate or Imprecise Result or Readings	12/04/2023	12/04/2023
					CT2406	Implantable	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	12/04/2023	12/04/2023
									A0908	Incorrect, Inadequate or Imprecise Result or Readings	12/04/2023	12/04/2023
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	12/04/2023	12/04/2023
									A0908	Incorrect, Inadequate or Imprecise Result or Readings	12/04/2023	12/04/2023
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	12/04/2023	12/04/2023
									A0908	Incorrect, Inadequate or Imprecise Result or Readings	12/04/2023	12/04/2023
					CT2407	Active implantable	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	12/04/2023	12/04/2023
									A0908	Incorrect, Inadequate or Imprecise Result or Readings	12/04/2023	12/04/2023
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	12/04/2023	12/04/2023
									A0908	Incorrect, Inadequate or Imprecise Result or Readings	12/04/2023	12/04/2023
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	12/04/2023	12/04/2023
									A0908	Incorrect, Inadequate or Imprecise Result or Readings	12/04/2023	12/04/2023
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	12/04/2023	12/04/2023
									A0908	Incorrect, Inadequate or Imprecise Result or Readings	12/04/2023	12/04/2023
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	12/04/2023	12/04/2023
									A0908	Incorrect, Inadequate or Imprecise Result or Readings	12/04/2023	12/04/2023
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	12/04/2023	12/04/2023
									A0908	Incorrect, Inadequate or Imprecise Result or Readings	12/04/2023	12/04/2023
					CT2727	Active implant devices/ancillaries	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	12/04/2023	12/04/2023
									A0908	Incorrect, Inadequate or Imprecise Result or Readings	12/04/2023	12/04/2023
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	12/04/2023	12/04/2023
									A0908	Incorrect, Inadequate or Imprecise Result or Readings	12/04/2023	12/04/2023
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	12/04/2023	12/04/2023
									A0908	Incorrect, Inadequate or Imprecise Result or Readings	12/04/2023	12/04/2023
					CT2750	Implantable cardiac electrical stimulators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	12/04/2023	12/04/2023
									A0908	Incorrect, Inadequate or Imprecise Result or Readings	12/04/2023	12/04/2023
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	12/04/2023	12/04/2023
									A0908	Incorrect, Inadequate or Imprecise Result or Readings	12/04/2023	12/04/2023
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	12/04/2023	12/04/2023
									A0908	Incorrect, Inadequate or Imprecise Result or Readings	12/04/2023	12/04/2023
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	12/04/2023	12/04/2023
									A0908	Incorrect, Inadequate or Imprecise Result or Readings	12/04/2023	12/04/2023
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	12/04/2023	12/04/2023
									A0908	Incorrect, Inadequate or Imprecise Result or Readings	12/04/2023	12/04/2023
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	12/04/2023	12/04/2023
									A0908	Incorrect, Inadequate or Imprecise Result or Readings	12/04/2023	12/04/2023

					CT2985	Active	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	12/04/2023	12/04/2023
									A0908	Incorrect, Inadequate or Imprecise Result or Readings	12/04/2023	12/04/2023
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	12/04/2023	12/04/2023
									A0908	Incorrect, Inadequate or Imprecise Result or Readings	12/04/2023	12/04/2023
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	12/04/2023	12/04/2023
									A0908	Incorrect, Inadequate or Imprecise Result or Readings	12/04/2023	12/04/2023
					CT3001	Implantable metal	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	12/04/2023	12/04/2023
									A0908	Incorrect, Inadequate or Imprecise Result or Readings	12/04/2023	12/04/2023
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	12/04/2023	12/04/2023
									A0908	Incorrect, Inadequate or Imprecise Result or Readings	12/04/2023	12/04/2023
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	12/04/2023	12/04/2023
									A0908	Incorrect, Inadequate or Imprecise Result or Readings	12/04/2023	12/04/2023
					CT3002	Implantable synthetic polymer	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	12/04/2023	12/04/2023
									A0908	Incorrect, Inadequate or Imprecise Result or Readings	12/04/2023	12/04/2023
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	12/04/2023	12/04/2023
									A0908	Incorrect, Inadequate or Imprecise Result or Readings	12/04/2023	12/04/2023
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	12/04/2023	12/04/2023
									A0908	Incorrect, Inadequate or Imprecise Result or Readings	12/04/2023	12/04/2023
					CT3006	Implantable inorganic material	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	12/04/2023	12/04/2023
									A0908	Incorrect, Inadequate or Imprecise Result or Readings	12/04/2023	12/04/2023
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	12/04/2023	12/04/2023
									A0908	Incorrect, Inadequate or Imprecise Result or Readings	12/04/2023	12/04/2023
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	12/04/2023	12/04/2023
									A0908	Incorrect, Inadequate or Imprecise Result or Readings	12/04/2023	12/04/2023
					CT314	Electrically powered	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	12/04/2023	12/04/2023
									A0908	Incorrect, Inadequate or Imprecise Result or Readings	12/04/2023	12/04/2023
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	12/04/2023	12/04/2023
									A0908	Incorrect, Inadequate or Imprecise Result or Readings	12/04/2023	12/04/2023
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	12/04/2023	12/04/2023
									A0908	Incorrect, Inadequate or Imprecise Result or Readings	12/04/2023	12/04/2023
					CT321	Long-term surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	12/04/2023	12/04/2023
									A0908	Incorrect, Inadequate or Imprecise Result or Readings	12/04/2023	12/04/2023
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	12/04/2023	12/04/2023
									A0908	Incorrect, Inadequate or Imprecise Result or Readings	12/04/2023	12/04/2023
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	12/04/2023	12/04/2023
									A0908	Incorrect, Inadequate or Imprecise Result or Readings	12/04/2023	12/04/2023
					CT449	Generators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	12/04/2023	12/04/2023
									A0908	Incorrect, Inadequate or Imprecise Result or Readings	12/04/2023	12/04/2023
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	12/04/2023	12/04/2023
									A0908	Incorrect, Inadequate or Imprecise Result or Readings	12/04/2023	12/04/2023
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	12/04/2023	12/04/2023
									A0908	Incorrect, Inadequate or Imprecise Result or Readings	12/04/2023	12/04/2023
					CT981	Single use	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	12/04/2023	12/04/2023
									A0908	Incorrect, Inadequate or Imprecise Result or Readings	12/04/2023	12/04/2023
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	12/04/2023	12/04/2023
									A0908	Incorrect, Inadequate or Imprecise Result or Readings	12/04/2023	12/04/2023
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	12/04/2023	12/04/2023
									A0908	Incorrect, Inadequate or Imprecise Result or Readings	12/04/2023	12/04/2023
					CT983	Surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	12/04/2023	12/04/2023
									A0908	Incorrect, Inadequate or Imprecise Result or Readings	12/04/2023	12/04/2023
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	12/04/2023	12/04/2023
									A0908	Incorrect, Inadequate or Imprecise Result or Readings	12/04/2023	12/04/2023
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	12/04/2023	12/04/2023
									A0908	Incorrect, Inadequate or Imprecise Result or Readings	12/04/2023	12/04/2023
12962	2023	All other reportable incidents	37265	Dual-chamber implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A072201	High Impedance		
							F26	No Health Consequences or Impact	A072201	High Impedance		
12963	2023	All other reportable incidents	47265	Dual-chamber implantable pacemaker, rate-responsive	CT141	Electrodes/Leads and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT145	Cardiovascular devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT233	Surgical	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT2406	Implantable	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		

							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT2407	Active implantable	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT2727	Active implant devices/ancillaries	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT2750	Implantable cardiac electrical stimulators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT2985	Active	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT3001	Implantable metal	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT3002	Implantable synthetic polymer	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT3006	Implantable inorganic material	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT314	Electrically powered	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT321	Long-term surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT449	Generators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT981	Single use	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT983	Surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
12964	2023	All other reportable incidents	47265	Dual-chamber implantable pacemaker, rate-responsive	CT141	Electrodes/Leads and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	17/04/2023	17/04/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	17/04/2023	17/04/2023
					CT145	Cardiovascular devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	17/04/2023	17/04/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	17/04/2023	17/04/2023
					CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	17/04/2023	17/04/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	17/04/2023	17/04/2023
					CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	17/04/2023	17/04/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	17/04/2023	17/04/2023
					CT233	Surgical	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	17/04/2023	17/04/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	17/04/2023	17/04/2023
					CT2406	Implantable	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	17/04/2023	17/04/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	17/04/2023	17/04/2023
					CT2407	Active implantable	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	17/04/2023	17/04/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	17/04/2023	17/04/2023
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	17/04/2023	17/04/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	17/04/2023	17/04/2023
					CT2727	Active implant devices/ancillaries	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	17/04/2023	17/04/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	17/04/2023	17/04/2023
					CT2750	Implantable cardiac electrical stimulators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	17/04/2023	17/04/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	17/04/2023	17/04/2023
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	17/04/2023	17/04/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	17/04/2023	17/04/2023
					CT2985	Active	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	17/04/2023	17/04/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	17/04/2023	17/04/2023
					CT3001	Implantable metal	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	17/04/2023	17/04/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	17/04/2023	17/04/2023
					CT3002	Implantable synthetic polymer	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	17/04/2023	17/04/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	17/04/2023	17/04/2023
					CT3006	Implantable inorganic material	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	17/04/2023	17/04/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	17/04/2023	17/04/2023
					CT314	Electrically powered	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	17/04/2023	17/04/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	17/04/2023	17/04/2023
					CT321	Long-term surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	17/04/2023	17/04/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	17/04/2023	17/04/2023
					CT449	Generators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	17/04/2023	17/04/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	17/04/2023	17/04/2023
					CT981	Single use	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	17/04/2023	17/04/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	17/04/2023	17/04/2023
					CT983	Surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	17/04/2023	17/04/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	17/04/2023	17/04/2023
12965	2023	All other reportable incidents	36241	Pacemaker/defibrillator lead	CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A040101	Fracture		
							F08	Hospitalization or Prolonged Hospitalization	A040101	Fracture		
							F1905	Device Revision or Replacement	A040101	Fracture		
12966	2023	All other reportable incidents	17882	Physiologic-monitoring defibrillation system	CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A150105	Key or Button Unresponsive/not Working		
							F27	Problem Identified During Non-Clinical Procedure	A150105	Key or Button Unresponsive/not Working		
12967	2023	All other reportable incidents	17882	Physiologic-monitoring defibrillation system	CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A1603	Failure of Device to Self-Test		
							F27	Problem Identified During Non-Clinical Procedure	A1603	Failure of Device to Self-Test		
12968	2023	All other reportable incidents	47270	Cardiac resynchronization therapy implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A160601	Premature Elective Replacement Indicator		
							F1903	Device Explantation	A160601	Premature Elective Replacement Indicator		
12969	2023	All other reportable incidents	35223	Endocardial/interventricular septal pacing lead	CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070909	Oversensing		
							F26	No Health Consequences or Impact	A070909	Oversensing		
12970	2023	All other reportable incidents	37265	Dual-chamber implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A1601	Device Alarm System		
							F26	No Health Consequences or Impact	A1601	Device Alarm System		
12971	2023		35223	Endocardial/interventricular septal pacing lead	CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A0401	Break		
									A070909	Oversensing		
							F23	Unexpected Medical Intervention	A0401	Break		
									A070909	Oversensing		
12972	2023	All other reportable incidents	60910	Coronary venous pacing lead	CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A0701	Capturing Problem		
									A070102	High Capture Threshold		

								A0722	Impedance Problem			
								A072201	High Impedance			
						F26	No Health Consequences or Impact	A0701	Capturing Problem			
								A070102	High Capture Threshold			
								A0722	Impedance Problem			
								A072201	High Impedance			
12973	2023	All other reportable incidents	35852	Single-chamber implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E2104	Electric Shock	A070909	Oversensing		
									A071302	Inappropriate/Inadequate Shock/Stimulation		
									A1301	Failure to Read Input Signal		
							F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A070909	Oversensing		
									A071302	Inappropriate/Inadequate Shock/Stimulation		
									A1301	Failure to Read Input Signal		
12974	2023				CT1684	Cardiac pacemakers and associated devices	E06	Heart	A07	Electrical /Electronic Property Problem		
									A22	Human-Device Interface Problem		
							E21	Procedural Complications	A07	Electrical /Electronic Property Problem		
									A22	Human-Device Interface Problem		
							E2403	No Clinical Signs, Symptoms or Conditions	A07	Electrical /Electronic Property Problem		
									A22	Human-Device Interface Problem		
							F17	Sedation	A07	Electrical /Electronic Property Problem		
									A22	Human-Device Interface Problem		
							F23	Unexpected Medical Intervention	A07	Electrical /Electronic Property Problem		
									A22	Human-Device Interface Problem		
							F26	No Health Consequences or Impact	A07	Electrical /Electronic Property Problem		
									A22	Human-Device Interface Problem		
12975	2023	All other reportable incidents	47265	Dual-chamber implantable pacemaker, rate-responsive	CT141	Electrodes/Leads and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	08/02/2023	08/02/2023
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	08/02/2023	08/02/2023
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	08/02/2023	08/02/2023
					CT145	Cardiovascular devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	08/02/2023	08/02/2023
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	08/02/2023	08/02/2023
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	08/02/2023	08/02/2023
					CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	08/02/2023	08/02/2023
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	08/02/2023	08/02/2023
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	08/02/2023	08/02/2023
					CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	08/02/2023	08/02/2023
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	08/02/2023	08/02/2023
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	08/02/2023	08/02/2023
					CT233	Surgical	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	08/02/2023	08/02/2023
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	08/02/2023	08/02/2023
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	08/02/2023	08/02/2023
					CT2406	Implantable	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	08/02/2023	08/02/2023
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	08/02/2023	08/02/2023
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	08/02/2023	08/02/2023
					CT2407	Active implantable	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	08/02/2023	08/02/2023
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	08/02/2023	08/02/2023
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	08/02/2023	08/02/2023
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	08/02/2023	08/02/2023
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	08/02/2023	08/02/2023
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	08/02/2023	08/02/2023
					CT2727	Active implant devices/ancillaries	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	08/02/2023	08/02/2023
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	08/02/2023	08/02/2023
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	08/02/2023	08/02/2023
					CT2750	Implantable cardiac electrical stimulators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	08/02/2023	08/02/2023
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	08/02/2023	08/02/2023
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	08/02/2023	08/02/2023
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	08/02/2023	08/02/2023
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	08/02/2023	08/02/2023
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	08/02/2023	08/02/2023
					CT2985	Active	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	08/02/2023	08/02/2023
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	08/02/2023	08/02/2023
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	08/02/2023	08/02/2023
					CT3001	Implantable metal	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	08/02/2023	08/02/2023
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	08/02/2023	08/02/2023
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	08/02/2023	08/02/2023
					CT3002	Implantable synthetic polymer	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	08/02/2023	08/02/2023
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	08/02/2023	08/02/2023
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	08/02/2023	08/02/2023
					CT3006	Implantable inorganic material	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	08/02/2023	08/02/2023
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	08/02/2023	08/02/2023
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	08/02/2023	08/02/2023
					CT314	Electrically powered	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	08/02/2023	08/02/2023
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	08/02/2023	08/02/2023
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	08/02/2023	08/02/2023
					CT321	Long-term surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	08/02/2023	08/02/2023
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	08/02/2023	08/02/2023
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	08/02/2023	08/02/2023
					CT449	Generators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	08/02/2023	08/02/2023
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	08/02/2023	08/02/2023
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	08/02/2023	08/02/2023
					CT981	Single use	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	08/02/2023	08/02/2023
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	08/02/2023	08/02/2023
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	08/02/2023	08/02/2023
					CT983	Surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	08/02/2023	08/02/2023
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	08/02/2023	08/02/2023
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	08/02/2023	08/02/2023
12976	2023		47265	Dual-chamber implantable pacemaker, rate-responsive	CT141	Electrodes/Leads and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	28/03/2023	28/03/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	28/03/2023	28/03/2023
					CT145	Cardiovascular devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	28/03/2023	28/03/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	28/03/2023	28/03/2023

					CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	28/03/2023	28/03/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	28/03/2023	28/03/2023
					CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	28/03/2023	28/03/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	28/03/2023	28/03/2023
					CT233	Surgical	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	28/03/2023	28/03/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	28/03/2023	28/03/2023
					CT2406	Implantable	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	28/03/2023	28/03/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	28/03/2023	28/03/2023
					CT2407	Active implantable	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	28/03/2023	28/03/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	28/03/2023	28/03/2023
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	28/03/2023	28/03/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	28/03/2023	28/03/2023
					CT2727	Active implant devices/ancillaries	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	28/03/2023	28/03/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	28/03/2023	28/03/2023
					CT2750	Implantable cardiac electrical stimulators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	28/03/2023	28/03/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	28/03/2023	28/03/2023
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	28/03/2023	28/03/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	28/03/2023	28/03/2023
					CT2985	Active	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	28/03/2023	28/03/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	28/03/2023	28/03/2023
					CT3001	Implantable metal	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	28/03/2023	28/03/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	28/03/2023	28/03/2023
					CT3002	Implantable synthetic polymer	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	28/03/2023	28/03/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	28/03/2023	28/03/2023
					CT3006	Implantable inorganic material	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	28/03/2023	28/03/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	28/03/2023	28/03/2023
					CT314	Electrically powered	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	28/03/2023	28/03/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	28/03/2023	28/03/2023
					CT321	Long-term surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	28/03/2023	28/03/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	28/03/2023	28/03/2023
					CT449	Generators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	28/03/2023	28/03/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	28/03/2023	28/03/2023
					CT981	Single use	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	28/03/2023	28/03/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	28/03/2023	28/03/2023
					CT983	Surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	28/03/2023	28/03/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	28/03/2023	28/03/2023
12977	2023	All other reportable incidents	35852	Single-chamber implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
12978	2023		47265	Dual-chamber implantable pacemaker, rate-responsive	CT141	Electrodes/Leads and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem		
							F1901	Additional Surgery	A24	No Apparent Device Problem		
					CT145	Cardiovascular devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem		
							F1901	Additional Surgery	A24	No Apparent Device Problem		
					CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem		
							F1901	Additional Surgery	A24	No Apparent Device Problem		
					CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem		
							F1901	Additional Surgery	A24	No Apparent Device Problem		
					CT233	Surgical	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem		
							F1901	Additional Surgery	A24	No Apparent Device Problem		
					CT2406	Implantable	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem		
							F1901	Additional Surgery	A24	No Apparent Device Problem		
					CT2407	Active implantable	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem		
							F1901	Additional Surgery	A24	No Apparent Device Problem		
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem		
							F1901	Additional Surgery	A24	No Apparent Device Problem		
					CT2727	Active implant devices/ancillaries	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem		
							F1901	Additional Surgery	A24	No Apparent Device Problem		
					CT2750	Implantable cardiac electrical stimulators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem		
							F1901	Additional Surgery	A24	No Apparent Device Problem		
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem		
							F1901	Additional Surgery	A24	No Apparent Device Problem		
					CT2985	Active	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem		
							F1901	Additional Surgery	A24	No Apparent Device Problem		
					CT3001	Implantable metal	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem		
							F1901	Additional Surgery	A24	No Apparent Device Problem		
					CT3002	Implantable synthetic polymer	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem		
							F1901	Additional Surgery	A24	No Apparent Device Problem		
					CT3006	Implantable inorganic material	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem		
							F1901	Additional Surgery	A24	No Apparent Device Problem		
					CT314	Electrically powered	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem		
							F1901	Additional Surgery	A24	No Apparent Device Problem		
					CT321	Long-term surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem		
							F1901	Additional Surgery	A24	No Apparent Device Problem		
					CT449	Generators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem		
							F1901	Additional Surgery	A24	No Apparent Device Problem		
					CT981	Single use	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem		
							F1901	Additional Surgery	A24	No Apparent Device Problem		
					CT983	Surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem		
							F1901	Additional Surgery	A24	No Apparent Device Problem		
12979	2023		47265	Dual-chamber implantable pacemaker, rate-responsive	CT141	Electrodes/Leads and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem		
							F1901	Additional Surgery	A24	No Apparent Device Problem		
					CT145	Cardiovascular devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem		
							F1901	Additional Surgery	A24	No Apparent Device Problem		
					CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem		
							F1901	Additional Surgery	A24	No Apparent Device Problem		
					CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem		
							F1901	Additional Surgery	A24	No Apparent Device Problem		
					CT233	Surgical	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem		
							F1901	Additional Surgery	A24	No Apparent Device Problem		

					CT2406	Implantable	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem		
							F1901	Additional Surgery	A24	No Apparent Device Problem		
					CT2407	Active implantable	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem		
							F1901	Additional Surgery	A24	No Apparent Device Problem		
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem		
							F1901	Additional Surgery	A24	No Apparent Device Problem		
					CT2727	Active implant devices/ancillaries	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem		
							F1901	Additional Surgery	A24	No Apparent Device Problem		
					CT2750	Implantable cardiac electrical stimulators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem		
							F1901	Additional Surgery	A24	No Apparent Device Problem		
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem		
							F1901	Additional Surgery	A24	No Apparent Device Problem		
					CT2985	Active	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem		
							F1901	Additional Surgery	A24	No Apparent Device Problem		
					CT3001	Implantable metal	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem		
							F1901	Additional Surgery	A24	No Apparent Device Problem		
					CT3002	Implantable synthetic polymer	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem		
							F1901	Additional Surgery	A24	No Apparent Device Problem		
					CT3006	Implantable inorganic material	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem		
							F1901	Additional Surgery	A24	No Apparent Device Problem		
					CT314	Electrically powered	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem		
							F1901	Additional Surgery	A24	No Apparent Device Problem		
					CT321	Long-term surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem		
							F1901	Additional Surgery	A24	No Apparent Device Problem		
					CT449	Generators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem		
							F1901	Additional Surgery	A24	No Apparent Device Problem		
					CT981	Single use	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem		
							F1901	Additional Surgery	A24	No Apparent Device Problem		
					CT983	Surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem		
							F1901	Additional Surgery	A24	No Apparent Device Problem		
12980	2023	All other reportable incidents	17882	Physiologic-monitoring defibrillation system	CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A0719	Unexpected Shutdown		
							F27	Problem Identified During Non-Clinical Procedure	A0719	Unexpected Shutdown		
12981	2023		35852	Single-chamber implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E2401	Insufficient Information	A07	Electrical /Electronic Property Problem		
									A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A07	Electrical /Electronic Property Problem		
									A070504	Premature Discharge of Battery		
12982	2023	All other reportable incidents	35223	Endocardial/interventricular septal pacing lead	CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070909	Oversensing		
							F22	Unexpected Diagnostic Intervention	A070909	Oversensing		
12983	2023	All other reportable incidents	36241	Pacemaker/defibrillator lead	CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070909	Oversensing		
							F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A070909	Oversensing		
12984	2023	All other reportable incidents	47263	Cardiac resynchronization therapy implantable pacemaker	CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A160601	Premature Elective Replacement Indicator		
							F1903	Device Explantation	A160601	Premature Elective Replacement Indicator		
					CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A160601	Premature Elective Replacement Indicator		
							F1903	Device Explantation	A160601	Premature Elective Replacement Indicator		
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A160601	Premature Elective Replacement Indicator		
							F1903	Device Explantation	A160601	Premature Elective Replacement Indicator		
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A160601	Premature Elective Replacement Indicator		
							F1903	Device Explantation	A160601	Premature Elective Replacement Indicator		
12985	2023	All other reportable incidents	17882	Physiologic-monitoring defibrillation system	CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A0709	Device Sensing Problem		
							F26	No Health Consequences or Impact	A0709	Device Sensing Problem		
12986	2023	All other reportable incidents	47265	Dual-chamber implantable pacemaker, rate-responsive	CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem		
							F1901	Additional Surgery	A24	No Apparent Device Problem		
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem		
							F1901	Additional Surgery	A24	No Apparent Device Problem		
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem		
							F1901	Additional Surgery	A24	No Apparent Device Problem		
12987	2023	All other reportable incidents	47265	Dual-chamber implantable pacemaker, rate-responsive	CT141	Electrodes/Leads and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070909	Oversensing		
									A090801	Signal Artifact/Noise		
							F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A070909	Oversensing		
									A090801	Signal Artifact/Noise		
					CT145	Cardiovascular devices	E2403	No Clinical Signs, Symptoms or Conditions	A070909	Oversensing		
									A090801	Signal Artifact/Noise		
							F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A070909	Oversensing		
									A090801	Signal Artifact/Noise		
					CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070909	Oversensing		
									A090801	Signal Artifact/Noise		
							F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A070909	Oversensing		
									A090801	Signal Artifact/Noise		
					CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A070909	Oversensing		
									A090801	Signal Artifact/Noise		
							F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A070909	Oversensing		
									A090801	Signal Artifact/Noise		
					CT233	Surgical	E2403	No Clinical Signs, Symptoms or Conditions	A070909	Oversensing		
									A090801	Signal Artifact/Noise		
							F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A070909	Oversensing		
									A090801	Signal Artifact/Noise		
					CT2406	Implantable	E2403	No Clinical Signs, Symptoms or Conditions	A070909	Oversensing		
									A090801	Signal Artifact/Noise		
							F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A070909	Oversensing		
									A090801	Signal Artifact/Noise		
					CT2407	Active implantable	E2403	No Clinical Signs, Symptoms or Conditions	A070909	Oversensing		
									A090801	Signal Artifact/Noise		
							F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A070909	Oversensing		
									A090801	Signal Artifact/Noise		
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A070909	Oversensing		
									A090801	Signal Artifact/Noise		
							F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A070909	Oversensing		
									A090801	Signal Artifact/Noise		

					CT2727	Active implant devices/ancillaries	E2403	No Clinical Signs, Symptoms or Conditions	A070909	Oversensing		
									A090801	Signal Artifact/Noise		
							F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A070909	Oversensing		
									A090801	Signal Artifact/Noise		
					CT2750	Implantable cardiac electrical stimulators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070909	Oversensing		
									A090801	Signal Artifact/Noise		
							F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A070909	Oversensing		
									A090801	Signal Artifact/Noise		
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A070909	Oversensing		
									A090801	Signal Artifact/Noise		
							F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A070909	Oversensing		
									A090801	Signal Artifact/Noise		
					CT2985	Active	E2403	No Clinical Signs, Symptoms or Conditions	A070909	Oversensing		
									A090801	Signal Artifact/Noise		
							F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A070909	Oversensing		
									A090801	Signal Artifact/Noise		
					CT3001	Implantable metal	E2403	No Clinical Signs, Symptoms or Conditions	A070909	Oversensing		
									A090801	Signal Artifact/Noise		
							F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A070909	Oversensing		
									A090801	Signal Artifact/Noise		
					CT3002	Implantable synthetic polymer	E2403	No Clinical Signs, Symptoms or Conditions	A070909	Oversensing		
									A090801	Signal Artifact/Noise		
							F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A070909	Oversensing		
									A090801	Signal Artifact/Noise		
					CT3006	Implantable inorganic material	E2403	No Clinical Signs, Symptoms or Conditions	A070909	Oversensing		
									A090801	Signal Artifact/Noise		
							F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A070909	Oversensing		
									A090801	Signal Artifact/Noise		
					CT314	Electrically powered	E2403	No Clinical Signs, Symptoms or Conditions	A070909	Oversensing		
									A090801	Signal Artifact/Noise		
							F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A070909	Oversensing		
									A090801	Signal Artifact/Noise		
					CT321	Long-term surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A070909	Oversensing		
									A090801	Signal Artifact/Noise		
							F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A070909	Oversensing		
									A090801	Signal Artifact/Noise		
					CT449	Generators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070909	Oversensing		
									A090801	Signal Artifact/Noise		
							F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A070909	Oversensing		
									A090801	Signal Artifact/Noise		
					CT981	Single use	E2403	No Clinical Signs, Symptoms or Conditions	A070909	Oversensing		
									A090801	Signal Artifact/Noise		
							F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A070909	Oversensing		
									A090801	Signal Artifact/Noise		
					CT983	Surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A070909	Oversensing		
									A090801	Signal Artifact/Noise		
							F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A070909	Oversensing		
									A090801	Signal Artifact/Noise		
12988	2023	All other reportable incidents	35852	Single-chamber implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A04	Material Integrity Problem	04/01/2023	04/01/2023
									A070909	Oversensing	04/01/2023	04/01/2023
									A0722	Impedance Problem	04/01/2023	04/01/2023
									A090801	Signal Artifact/Noise	04/01/2023	04/01/2023
									A150207	Difficult to Remove	04/01/2023	04/01/2023
							F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A04	Material Integrity Problem	04/01/2023	04/01/2023
									A070909	Oversensing	04/01/2023	04/01/2023
									A0722	Impedance Problem	04/01/2023	04/01/2023
									A090801	Signal Artifact/Noise	04/01/2023	04/01/2023
									A150207	Difficult to Remove	04/01/2023	04/01/2023
12989	2023	All other reportable incidents	35852	Single-chamber implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E2104	Electric Shock	A070909	Oversensing		
									A071302	Inappropriate/Inadequate Shock/Stimulation		
									A1301	Failure to Read Input Signal		
							F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A070909	Oversensing		
									A071302	Inappropriate/Inadequate Shock/Stimulation		
									A1301	Failure to Read Input Signal		
12990	2023		47265	Dual-chamber implantable pacemaker, rate-responsive	CT141	Electrodes/Leads and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	14/04/2023	14/04/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	14/04/2023	14/04/2023
					CT145	Cardiovascular devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	14/04/2023	14/04/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	14/04/2023	14/04/2023
					CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	14/04/2023	14/04/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	14/04/2023	14/04/2023
					CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	14/04/2023	14/04/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	14/04/2023	14/04/2023
					CT233	Surgical	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	14/04/2023	14/04/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	14/04/2023	14/04/2023
					CT2406	Implantable	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	14/04/2023	14/04/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	14/04/2023	14/04/2023
					CT2407	Active implantable	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	14/04/2023	14/04/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	14/04/2023	14/04/2023
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	14/04/2023	14/04/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	14/04/2023	14/04/2023
					CT2727	Active implant devices/ancillaries	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	14/04/2023	14/04/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	14/04/2023	14/04/2023
					CT2750	Implantable cardiac electrical stimulators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	14/04/2023	14/04/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	14/04/2023	14/04/2023
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	14/04/2023	14/04/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	14/04/2023	14/04/2023
					CT2985	Active	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	14/04/2023	14/04/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	14/04/2023	14/04/2023

					CT3001	Implantable metal	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	14/04/2023	14/04/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	14/04/2023	14/04/2023
					CT3002	Implantable synthetic polymer	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	14/04/2023	14/04/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	14/04/2023	14/04/2023
					CT3006	Implantable inorganic material	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	14/04/2023	14/04/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	14/04/2023	14/04/2023
					CT314	Electrically powered	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	14/04/2023	14/04/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	14/04/2023	14/04/2023
					CT321	Long-term surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	14/04/2023	14/04/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	14/04/2023	14/04/2023
					CT449	Generators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	14/04/2023	14/04/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	14/04/2023	14/04/2023
					CT981	Single use	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	14/04/2023	14/04/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	14/04/2023	14/04/2023
					CT983	Surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	14/04/2023	14/04/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	14/04/2023	14/04/2023
12991	2023	All other reportable incidents	17882	Physiologic-monitoring defibrillation system	CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A071301	Failure to Deliver Shock/Stimulation		
							F26	No Health Consequences or Impact	A071301	Failure to Deliver Shock/Stimulation		
12992	2023		47265	Dual-chamber implantable pacemaker, rate-responsive	CT141	Electrodes/Leads and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	14/04/2023	14/04/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	14/04/2023	14/04/2023
					CT145	Cardiovascular devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	14/04/2023	14/04/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	14/04/2023	14/04/2023
					CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	14/04/2023	14/04/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	14/04/2023	14/04/2023
					CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	14/04/2023	14/04/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	14/04/2023	14/04/2023
					CT233	Surgical	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	14/04/2023	14/04/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	14/04/2023	14/04/2023
					CT2406	Implantable	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	14/04/2023	14/04/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	14/04/2023	14/04/2023
					CT2407	Active implantable	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	14/04/2023	14/04/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	14/04/2023	14/04/2023
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	14/04/2023	14/04/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	14/04/2023	14/04/2023
					CT2727	Active implant devices/ancillaries	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	14/04/2023	14/04/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	14/04/2023	14/04/2023
					CT2750	Implantable cardiac electrical stimulators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	14/04/2023	14/04/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	14/04/2023	14/04/2023
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	14/04/2023	14/04/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	14/04/2023	14/04/2023
					CT2985	Active	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	14/04/2023	14/04/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	14/04/2023	14/04/2023
					CT3001	Implantable metal	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	14/04/2023	14/04/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	14/04/2023	14/04/2023
					CT3002	Implantable synthetic polymer	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	14/04/2023	14/04/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	14/04/2023	14/04/2023
					CT3006	Implantable inorganic material	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	14/04/2023	14/04/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	14/04/2023	14/04/2023
					CT314	Electrically powered	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	14/04/2023	14/04/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	14/04/2023	14/04/2023
					CT321	Long-term surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	14/04/2023	14/04/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	14/04/2023	14/04/2023
					CT449	Generators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	14/04/2023	14/04/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	14/04/2023	14/04/2023
					CT981	Single use	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	14/04/2023	14/04/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	14/04/2023	14/04/2023
					CT983	Surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	14/04/2023	14/04/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	14/04/2023	14/04/2023
12993	2023		47265	Dual-chamber implantable pacemaker, rate-responsive	CT141	Electrodes/Leads and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	20/04/2023	20/04/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	20/04/2023	20/04/2023
					CT145	Cardiovascular devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	20/04/2023	20/04/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	20/04/2023	20/04/2023
					CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	20/04/2023	20/04/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	20/04/2023	20/04/2023
					CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	20/04/2023	20/04/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	20/04/2023	20/04/2023
					CT233	Surgical	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	20/04/2023	20/04/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	20/04/2023	20/04/2023
					CT2406	Implantable	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	20/04/2023	20/04/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	20/04/2023	20/04/2023
					CT2407	Active implantable	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	20/04/2023	20/04/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	20/04/2023	20/04/2023
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	20/04/2023	20/04/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	20/04/2023	20/04/2023
					CT2727	Active implant devices/ancillaries	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	20/04/2023	20/04/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	20/04/2023	20/04/2023
					CT2750	Implantable cardiac electrical stimulators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	20/04/2023	20/04/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	20/04/2023	20/04/2023
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	20/04/2023	20/04/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	20/04/2023	20/04/2023
					CT2985	Active	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	20/04/2023	20/04/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	20/04/2023	20/04/2023
					CT3001	Implantable metal	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	20/04/2023	20/04/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	20/04/2023	20/04/2023
					CT3002	Implantable synthetic polymer	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	20/04/2023	20/04/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	20/04/2023	20/04/2023
					CT3006	Implantable inorganic material	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	20/04/2023	20/04/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	20/04/2023	20/04/2023

					CT314	Electrically powered	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	20/04/2023	20/04/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	20/04/2023	20/04/2023
					CT321	Long-term surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	20/04/2023	20/04/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	20/04/2023	20/04/2023
					CT449	Generators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	20/04/2023	20/04/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	20/04/2023	20/04/2023
					CT981	Single use	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	20/04/2023	20/04/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	20/04/2023	20/04/2023
					CT983	Surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	20/04/2023	20/04/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	20/04/2023	20/04/2023
12994	2023	All other reportable incidents	35223	Endocardial/interventricular septal pacing lead	CT1684	Cardiac pacemakers and associated devices	E2402	Appropriate Clinical Signs, Symptoms and Conditions Term/Code not Available	A040101	Fracture		
									A070102	High Capture Threshold		
									A0712	Pacing Problem		
									A072201	High Impedance		
									A090801	Signal Artifact/Noise		
							F08	Hospitalization or Prolonged Hospitalization	A040101	Fracture		
									A070102	High Capture Threshold		
									A0712	Pacing Problem		
									A072201	High Impedance		
									A090801	Signal Artifact/Noise		
							F12	Serious Injury/ Illness/ Impairment	A040101	Fracture		
									A070102	High Capture Threshold		
									A0712	Pacing Problem		
									A072201	High Impedance		
									A090801	Signal Artifact/Noise		
							F23	Unexpected Medical Intervention	A040101	Fracture		
									A070102	High Capture Threshold		
									A0712	Pacing Problem		
									A072201	High Impedance		
									A090801	Signal Artifact/Noise		
12995	2023		47265	Dual-chamber implantable pacemaker, rate-responsive	CT141	Electrodes/Leads and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	04/02/2023	04/02/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	04/02/2023	04/02/2023
					CT145	Cardiovascular devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	04/02/2023	04/02/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	04/02/2023	04/02/2023
					CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	04/02/2023	04/02/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	04/02/2023	04/02/2023
					CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	04/02/2023	04/02/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	04/02/2023	04/02/2023
					CT233	Surgical	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	04/02/2023	04/02/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	04/02/2023	04/02/2023
					CT2406	Implantable	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	04/02/2023	04/02/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	04/02/2023	04/02/2023
					CT2407	Active implantable	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	04/02/2023	04/02/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	04/02/2023	04/02/2023
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	04/02/2023	04/02/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	04/02/2023	04/02/2023
					CT2727	Active implant devices/ancillaries	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	04/02/2023	04/02/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	04/02/2023	04/02/2023
					CT2750	Implantable cardiac electrical stimulators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	04/02/2023	04/02/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	04/02/2023	04/02/2023
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	04/02/2023	04/02/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	04/02/2023	04/02/2023
					CT2985	Active	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	04/02/2023	04/02/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	04/02/2023	04/02/2023
					CT3001	Implantable metal	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	04/02/2023	04/02/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	04/02/2023	04/02/2023
					CT3002	Implantable synthetic polymer	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	04/02/2023	04/02/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	04/02/2023	04/02/2023
					CT3006	Implantable inorganic material	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	04/02/2023	04/02/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	04/02/2023	04/02/2023
					CT314	Electrically powered	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	04/02/2023	04/02/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	04/02/2023	04/02/2023
					CT321	Long-term surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	04/02/2023	04/02/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	04/02/2023	04/02/2023
					CT449	Generators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	04/02/2023	04/02/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	04/02/2023	04/02/2023
					CT981	Single use	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	04/02/2023	04/02/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	04/02/2023	04/02/2023
					CT983	Surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	04/02/2023	04/02/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	04/02/2023	04/02/2023
12996	2023		47265	Dual-chamber implantable pacemaker, rate-responsive	CT141	Electrodes/Leads and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	04/02/2023	04/02/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	04/02/2023	04/02/2023
					CT145	Cardiovascular devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	04/02/2023	04/02/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	04/02/2023	04/02/2023
					CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	04/02/2023	04/02/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	04/02/2023	04/02/2023
					CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	04/02/2023	04/02/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	04/02/2023	04/02/2023
					CT233	Surgical	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	04/02/2023	04/02/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	04/02/2023	04/02/2023
					CT2406	Implantable	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	04/02/2023	04/02/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	04/02/2023	04/02/2023
					CT2407	Active implantable	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	04/02/2023	04/02/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	04/02/2023	04/02/2023
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	04/02/2023	04/02/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	04/02/2023	04/02/2023
					CT2727	Active implant devices/ancillaries	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	04/02/2023	04/02/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	04/02/2023	04/02/2023

					CT2750	Implantable cardiac electrical stimulators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	04/02/2023	04/02/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	04/02/2023	04/02/2023
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	04/02/2023	04/02/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	04/02/2023	04/02/2023
					CT2985	Active	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	04/02/2023	04/02/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	04/02/2023	04/02/2023
					CT3001	Implantable metal	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	04/02/2023	04/02/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	04/02/2023	04/02/2023
					CT3002	Implantable synthetic polymer	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	04/02/2023	04/02/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	04/02/2023	04/02/2023
					CT3006	Implantable inorganic material	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	04/02/2023	04/02/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	04/02/2023	04/02/2023
					CT314	Electrically powered	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	04/02/2023	04/02/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	04/02/2023	04/02/2023
					CT321	Long-term surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	04/02/2023	04/02/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	04/02/2023	04/02/2023
					CT449	Generators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	04/02/2023	04/02/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	04/02/2023	04/02/2023
					CT981	Single use	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	04/02/2023	04/02/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	04/02/2023	04/02/2023
					CT983	Surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	04/02/2023	04/02/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	04/02/2023	04/02/2023
12997	2023		47265	Dual-chamber implantable pacemaker, rate-responsive	CT141	Electrodes/Leads and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem		
							F1901	Additional Surgery	A24	No Apparent Device Problem		
					CT145	Cardiovascular devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem		
							F1901	Additional Surgery	A24	No Apparent Device Problem		
					CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem		
							F1901	Additional Surgery	A24	No Apparent Device Problem		
					CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem		
							F1901	Additional Surgery	A24	No Apparent Device Problem		
					CT233	Surgical	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem		
							F1901	Additional Surgery	A24	No Apparent Device Problem		
					CT2406	Implantable	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem		
							F1901	Additional Surgery	A24	No Apparent Device Problem		
					CT2407	Active implantable	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem		
							F1901	Additional Surgery	A24	No Apparent Device Problem		
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem		
							F1901	Additional Surgery	A24	No Apparent Device Problem		
					CT2727	Active implant devices/ancillaries	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem		
							F1901	Additional Surgery	A24	No Apparent Device Problem		
					CT2750	Implantable cardiac electrical stimulators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem		
							F1901	Additional Surgery	A24	No Apparent Device Problem		
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem		
							F1901	Additional Surgery	A24	No Apparent Device Problem		
					CT2985	Active	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem		
							F1901	Additional Surgery	A24	No Apparent Device Problem		
					CT3001	Implantable metal	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem		
							F1901	Additional Surgery	A24	No Apparent Device Problem		
					CT3002	Implantable synthetic polymer	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem		
							F1901	Additional Surgery	A24	No Apparent Device Problem		
					CT3006	Implantable inorganic material	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem		
							F1901	Additional Surgery	A24	No Apparent Device Problem		
					CT314	Electrically powered	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem		
							F1901	Additional Surgery	A24	No Apparent Device Problem		
					CT321	Long-term surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem		
							F1901	Additional Surgery	A24	No Apparent Device Problem		
					CT449	Generators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem		
							F1901	Additional Surgery	A24	No Apparent Device Problem		
					CT981	Single use	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem		
							F1901	Additional Surgery	A24	No Apparent Device Problem		
					CT983	Surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem		
							F1901	Additional Surgery	A24	No Apparent Device Problem		
12998	2023		47265	Dual-chamber implantable pacemaker, rate-responsive	CT141	Electrodes/Leads and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem		
							F1901	Additional Surgery	A24	No Apparent Device Problem		
					CT145	Cardiovascular devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem		
							F1901	Additional Surgery	A24	No Apparent Device Problem		
					CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem		
							F1901	Additional Surgery	A24	No Apparent Device Problem		
					CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem		
							F1901	Additional Surgery	A24	No Apparent Device Problem		
					CT233	Surgical	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem		
							F1901	Additional Surgery	A24	No Apparent Device Problem		
					CT2406	Implantable	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem		
							F1901	Additional Surgery	A24	No Apparent Device Problem		
					CT2407	Active implantable	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem		
							F1901	Additional Surgery	A24	No Apparent Device Problem		
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem		
							F1901	Additional Surgery	A24	No Apparent Device Problem		
					CT2727	Active implant devices/ancillaries	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem		
							F1901	Additional Surgery	A24	No Apparent Device Problem		
					CT2750	Implantable cardiac electrical stimulators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem		
							F1901	Additional Surgery	A24	No Apparent Device Problem		
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem		
							F1901	Additional Surgery	A24	No Apparent Device Problem		
					CT2985	Active	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem		
							F1901	Additional Surgery	A24	No Apparent Device Problem		
					CT3001	Implantable metal	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem		
							F1901	Additional Surgery	A24	No Apparent Device Problem		

					CT3002	Implantable synthetic polymer	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem		
							F1901	Additional Surgery	A24	No Apparent Device Problem		
					CT3006	Implantable inorganic material	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem		
							F1901	Additional Surgery	A24	No Apparent Device Problem		
					CT314	Electrically powered	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem		
							F1901	Additional Surgery	A24	No Apparent Device Problem		
					CT321	Long-term surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem		
							F1901	Additional Surgery	A24	No Apparent Device Problem		
					CT449	Generators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem		
							F1901	Additional Surgery	A24	No Apparent Device Problem		
					CT981	Single use	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem		
							F1901	Additional Surgery	A24	No Apparent Device Problem		
					CT983	Surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem		
							F1901	Additional Surgery	A24	No Apparent Device Problem		
12999	2023		47265	Dual-chamber implantable pacemaker, rate-responsive	CT141	Electrodes/Leads and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	05/02/2023	05/02/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	05/02/2023	05/02/2023
					CT145	Cardiovascular devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	05/02/2023	05/02/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	05/02/2023	05/02/2023
					CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	05/02/2023	05/02/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	05/02/2023	05/02/2023
					CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	05/02/2023	05/02/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	05/02/2023	05/02/2023
					CT233	Surgical	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	05/02/2023	05/02/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	05/02/2023	05/02/2023
					CT2406	Implantable	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	05/02/2023	05/02/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	05/02/2023	05/02/2023
					CT2407	Active implantable	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	05/02/2023	05/02/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	05/02/2023	05/02/2023
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	05/02/2023	05/02/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	05/02/2023	05/02/2023
					CT2727	Active implant devices/ancillaries	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	05/02/2023	05/02/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	05/02/2023	05/02/2023
					CT2750	Implantable cardiac electrical stimulators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	05/02/2023	05/02/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	05/02/2023	05/02/2023
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	05/02/2023	05/02/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	05/02/2023	05/02/2023
					CT2985	Active	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	05/02/2023	05/02/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	05/02/2023	05/02/2023
					CT3001	Implantable metal	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	05/02/2023	05/02/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	05/02/2023	05/02/2023
					CT3002	Implantable synthetic polymer	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	05/02/2023	05/02/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	05/02/2023	05/02/2023
					CT3006	Implantable inorganic material	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	05/02/2023	05/02/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	05/02/2023	05/02/2023
					CT314	Electrically powered	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	05/02/2023	05/02/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	05/02/2023	05/02/2023
					CT321	Long-term surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	05/02/2023	05/02/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	05/02/2023	05/02/2023
					CT449	Generators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	05/02/2023	05/02/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	05/02/2023	05/02/2023
					CT981	Single use	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	05/02/2023	05/02/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	05/02/2023	05/02/2023
					CT983	Surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	05/02/2023	05/02/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	05/02/2023	05/02/2023
13000	2023	All other reportable incidents	47263	Cardiac resynchronization therapy implantable pacemaker	CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	12/01/2023	12/01/2023
							F08	Hospitalization or Prolonged Hospitalization	A24	No Apparent Device Problem	12/01/2023	12/01/2023
							F1905	Device Revision or Replacement	A24	No Apparent Device Problem	12/01/2023	12/01/2023
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	12/01/2023	12/01/2023
							F08	Hospitalization or Prolonged Hospitalization	A24	No Apparent Device Problem	12/01/2023	12/01/2023
							F1905	Device Revision or Replacement	A24	No Apparent Device Problem	12/01/2023	12/01/2023
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	12/01/2023	12/01/2023
							F08	Hospitalization or Prolonged Hospitalization	A24	No Apparent Device Problem	12/01/2023	12/01/2023
							F1905	Device Revision or Replacement	A24	No Apparent Device Problem	12/01/2023	12/01/2023
13001	2023	All other reportable incidents	47267	Single-chamber implantable pacemaker, rate-responsive	CT141	Electrodes/Leads and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
					CT145	Cardiovascular devices	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
					CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
					CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
					CT233	Surgical	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
					CT2406	Implantable	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
					CT2407	Active implantable	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem		

									A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
					CT2727	Active implant devices/ancillaries	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
					CT2750	Implantable cardiac electrical stimulators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
					CT2985	Active	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
					CT3001	Implantable metal	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
					CT3002	Implantable synthetic polymer	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
					CT3006	Implantable inorganic material	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
					CT314	Electrically powered	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
					CT321	Long-term surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
					CT449	Generators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
					CT981	Single use	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
					CT983	Surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
13002	2023	All other reportable incidents	47265	Dual-chamber implantable pacemaker, rate-responsive	CT141	Electrodes/Leads and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
							F24	Insufficient Health Impact Information	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
					CT145	Cardiovascular devices	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
							F24	Insufficient Health Impact Information	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
					CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
							F24	Insufficient Health Impact Information	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
					CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
							F24	Insufficient Health Impact Information	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
					CT233	Surgical	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
							F24	Insufficient Health Impact Information	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
					CT2406	Implantable	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
							F24	Insufficient Health Impact Information	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
					CT2407	Active implantable	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
							F24	Insufficient Health Impact Information	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
							F24	Insufficient Health Impact Information	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
					CT2727	Active implant devices/ancillaries	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem		

									A070504	Premature Discharge of Battery		
							F24	Insufficient Health Impact Information	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
					CT2750	Implantable cardiac electrical stimulators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
							F24	Insufficient Health Impact Information	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
							F24	Insufficient Health Impact Information	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
					CT2985	Active	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
							F24	Insufficient Health Impact Information	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
					CT3001	Implantable metal	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
							F24	Insufficient Health Impact Information	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
					CT3002	Implantable synthetic polymer	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
							F24	Insufficient Health Impact Information	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
					CT3006	Implantable inorganic material	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
							F24	Insufficient Health Impact Information	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
					CT314	Electrically powered	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
							F24	Insufficient Health Impact Information	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
					CT321	Long-term surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
							F24	Insufficient Health Impact Information	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
					CT449	Generators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
							F24	Insufficient Health Impact Information	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
					CT981	Single use	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
							F24	Insufficient Health Impact Information	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
					CT983	Surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
							F24	Insufficient Health Impact Information	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
13003	2023	All other reportable incidents	36241	Pacemaker/defibrillator lead	CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A090801	Signal Artifact/Noise	19/04/2023	19/04/2023
							F1908	Prolonged Surgery	A090801	Signal Artifact/Noise	19/04/2023	19/04/2023
13004	2023	All other reportable incidents	47265	Dual-chamber implantable pacemaker, rate-responsive	CT141	Electrodes/Leads and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT145	Cardiovascular devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT233	Surgical	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT2406	Implantable	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT2407	Active implantable	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT2727	Active implant devices/ancillaries	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT2750	Implantable cardiac electrical stimulators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT2985	Active	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT3001	Implantable metal	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT3002	Implantable synthetic polymer	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT3006	Implantable inorganic material	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT314	Electrically powered	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT321	Long-term surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT449	Generators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT981	Single use	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT983	Surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		

							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
13005	2023	All other reportable incidents	36241	Pacemaker/defibrillator lead	CT1684	Cardiac pacemakers and associated devices	E2103	Device Overstimulation of Tissue	A051201	Device Dislodged or Dislocated		
									A070101	Failure to Capture		
									A072201	High Impedance		
							F08	Hospitalization or Prolonged Hospitalization	A051201	Device Dislodged or Dislocated		
									A070101	Failure to Capture		
									A072201	High Impedance		
							F1905	Device Revision or Replacement	A051201	Device Dislodged or Dislocated		
									A070101	Failure to Capture		
									A072201	High Impedance		
13006	2023	All other reportable incidents	47265	Dual-chamber implantable pacemaker, rate-responsive	CT141	Electrodes/Leads and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT145	Cardiovascular devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT233	Surgical	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT2406	Implantable	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT2407	Active implantable	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT2727	Active implant devices/ancillaries	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT2750	Implantable cardiac electrical stimulators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT2985	Active	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT3001	Implantable metal	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT3002	Implantable synthetic polymer	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT3006	Implantable inorganic material	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT314	Electrically powered	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT321	Long-term surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT449	Generators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT981	Single use	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT983	Surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
13007	2023	All other reportable incidents	35852	Single-chamber implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
13008	2023	All other reportable incidents	47265	Dual-chamber implantable pacemaker, rate-responsive	CT141	Electrodes/Leads and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	28/03/2023	28/03/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	28/03/2023	28/03/2023
					CT145	Cardiovascular devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	28/03/2023	28/03/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	28/03/2023	28/03/2023
					CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	28/03/2023	28/03/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	28/03/2023	28/03/2023
					CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	28/03/2023	28/03/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	28/03/2023	28/03/2023
					CT233	Surgical	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	28/03/2023	28/03/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	28/03/2023	28/03/2023
					CT2406	Implantable	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	28/03/2023	28/03/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	28/03/2023	28/03/2023
					CT2407	Active implantable	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	28/03/2023	28/03/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	28/03/2023	28/03/2023
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	28/03/2023	28/03/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	28/03/2023	28/03/2023
					CT2727	Active implant devices/ancillaries	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	28/03/2023	28/03/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	28/03/2023	28/03/2023
					CT2750	Implantable cardiac electrical stimulators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	28/03/2023	28/03/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	28/03/2023	28/03/2023
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	28/03/2023	28/03/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	28/03/2023	28/03/2023
					CT2985	Active	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	28/03/2023	28/03/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	28/03/2023	28/03/2023
					CT3001	Implantable metal	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	28/03/2023	28/03/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	28/03/2023	28/03/2023
					CT3002	Implantable synthetic polymer	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	28/03/2023	28/03/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	28/03/2023	28/03/2023
					CT3006	Implantable inorganic material	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	28/03/2023	28/03/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	28/03/2023	28/03/2023
					CT314	Electrically powered	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	28/03/2023	28/03/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	28/03/2023	28/03/2023
					CT321	Long-term surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	28/03/2023	28/03/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	28/03/2023	28/03/2023

					CT449	Generators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	28/03/2023	28/03/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	28/03/2023	28/03/2023
					CT981	Single use	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	28/03/2023	28/03/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	28/03/2023	28/03/2023
					CT983	Surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	28/03/2023	28/03/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	28/03/2023	28/03/2023
13009	2023	All other reportable incidents	35852	Single-chamber implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E2104	Electric Shock	A070909	Oversensing		
									A071302	Inappropriate/Inadequate Shock/Stimulation		
									A1301	Failure to Read Input Signal		
							F08	Hospitalization or Prolonged Hospitalization	A070909	Oversensing		
									A071302	Inappropriate/Inadequate Shock/Stimulation		
									A1301	Failure to Read Input Signal		
							F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A070909	Oversensing		
									A071302	Inappropriate/Inadequate Shock/Stimulation		
									A1301	Failure to Read Input Signal		
13010	2023	All other reportable incidents	17882	Physiologic-monitoring defibrillation system	CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A051203	Device Fell		
							F27	Problem Identified During Non-Clinical Procedure	A051203	Device Fell		
13011	2023	All other reportable incidents	35852	Single-chamber implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E2104	Electric Shock	A070909	Oversensing		
									A071302	Inappropriate/Inadequate Shock/Stimulation		
									A1301	Failure to Read Input Signal		
							F08	Hospitalization or Prolonged Hospitalization	A070909	Oversensing		
									A071302	Inappropriate/Inadequate Shock/Stimulation		
									A1301	Failure to Read Input Signal		
							F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A070909	Oversensing		
									A071302	Inappropriate/Inadequate Shock/Stimulation		
									A1301	Failure to Read Input Signal		
13012	2023	All other reportable incidents	35852	Single-chamber implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E2104	Electric Shock	A070909	Oversensing		
									A071302	Inappropriate/Inadequate Shock/Stimulation		
							F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A070909	Oversensing		
									A071302	Inappropriate/Inadequate Shock/Stimulation		
13013	2023	All other reportable incidents	35223	Endocardial/interventricular septal pacing lead	CT1684	Cardiac pacemakers and associated devices	E0112	Dizziness	A070909	Oversensing		
									A0712	Pacing Problem		
							E011903	Syncope/Fainting	A070909	Oversensing		
									A0712	Pacing Problem		
							E0601	Arrhythmia	A070909	Oversensing		
									A0712	Pacing Problem		
							F1203	Life Threatening Illness or Injury	A070909	Oversensing		
									A0712	Pacing Problem		
13014	2023		47265	Dual-chamber implantable pacemaker, rate-responsive	CT141	Electrodes/Leads and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	20/04/2023	20/04/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	20/04/2023	20/04/2023
					CT145	Cardiovascular devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	20/04/2023	20/04/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	20/04/2023	20/04/2023
					CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	20/04/2023	20/04/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	20/04/2023	20/04/2023
					CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	20/04/2023	20/04/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	20/04/2023	20/04/2023
					CT233	Surgical	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	20/04/2023	20/04/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	20/04/2023	20/04/2023
					CT2406	Implantable	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	20/04/2023	20/04/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	20/04/2023	20/04/2023
					CT2407	Active implantable	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	20/04/2023	20/04/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	20/04/2023	20/04/2023
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	20/04/2023	20/04/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	20/04/2023	20/04/2023
					CT2727	Active implant devices/ancillaries	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	20/04/2023	20/04/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	20/04/2023	20/04/2023
					CT2750	Implantable cardiac electrical stimulators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	20/04/2023	20/04/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	20/04/2023	20/04/2023
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	20/04/2023	20/04/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	20/04/2023	20/04/2023
					CT2985	Active	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	20/04/2023	20/04/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	20/04/2023	20/04/2023
					CT3001	Implantable metal	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	20/04/2023	20/04/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	20/04/2023	20/04/2023
					CT3002	Implantable synthetic polymer	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	20/04/2023	20/04/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	20/04/2023	20/04/2023
					CT3006	Implantable inorganic material	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	20/04/2023	20/04/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	20/04/2023	20/04/2023
					CT314	Electrically powered	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	20/04/2023	20/04/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	20/04/2023	20/04/2023
					CT321	Long-term surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	20/04/2023	20/04/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	20/04/2023	20/04/2023
					CT449	Generators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	20/04/2023	20/04/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	20/04/2023	20/04/2023
					CT981	Single use	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	20/04/2023	20/04/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	20/04/2023	20/04/2023
					CT983	Surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	20/04/2023	20/04/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	20/04/2023	20/04/2023
13015	2023	All other reportable incidents	35223	Endocardial/interventricular septal pacing lead	CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A0722	Impedance Problem		
									A072202	Low Impedance		
									A090801	Signal Artifact/Noise		
							F23	Unexpected Medical Intervention	A0722	Impedance Problem		
									A072202	Low Impedance		
									A090801	Signal Artifact/Noise		
13016	2023	All other reportable incidents	35223	Endocardial/interventricular septal pacing lead	CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A0401	Break		
									A070909	Oversensing		
							F22	Unexpected Diagnostic Intervention	A0401	Break		
									A070909	Oversensing		

13017	2023		47265	Dual-chamber implantable pacemaker, rate-responsive	CT141	Electrodes/Leads and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	16/03/2023	16/03/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	16/03/2023	16/03/2023
					CT145	Cardiovascular devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	16/03/2023	16/03/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	16/03/2023	16/03/2023
					CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	16/03/2023	16/03/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	16/03/2023	16/03/2023
					CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	16/03/2023	16/03/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	16/03/2023	16/03/2023
					CT233	Surgical	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	16/03/2023	16/03/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	16/03/2023	16/03/2023
					CT2406	Implantable	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	16/03/2023	16/03/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	16/03/2023	16/03/2023
					CT2407	Active implantable	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	16/03/2023	16/03/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	16/03/2023	16/03/2023
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	16/03/2023	16/03/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	16/03/2023	16/03/2023
					CT2727	Active implant devices/ancillaries	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	16/03/2023	16/03/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	16/03/2023	16/03/2023
					CT2750	Implantable cardiac electrical stimulators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	16/03/2023	16/03/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	16/03/2023	16/03/2023
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	16/03/2023	16/03/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	16/03/2023	16/03/2023
					CT2985	Active	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	16/03/2023	16/03/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	16/03/2023	16/03/2023
					CT3001	Implantable metal	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	16/03/2023	16/03/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	16/03/2023	16/03/2023
					CT3002	Implantable synthetic polymer	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	16/03/2023	16/03/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	16/03/2023	16/03/2023
					CT3006	Implantable inorganic material	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	16/03/2023	16/03/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	16/03/2023	16/03/2023
					CT314	Electrically powered	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	16/03/2023	16/03/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	16/03/2023	16/03/2023
					CT321	Long-term surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	16/03/2023	16/03/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	16/03/2023	16/03/2023
					CT449	Generators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	16/03/2023	16/03/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	16/03/2023	16/03/2023
					CT981	Single use	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	16/03/2023	16/03/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	16/03/2023	16/03/2023
					CT983	Surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	16/03/2023	16/03/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	16/03/2023	16/03/2023
13018	2023	All other reportable incidents	47265	Dual-chamber implantable pacemaker, rate-responsive	CT141	Electrodes/Leads and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT145	Cardiovascular devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT233	Surgical	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT2406	Implantable	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT2407	Active implantable	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT2727	Active implant devices/ancillaries	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT2750	Implantable cardiac electrical stimulators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT2985	Active	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT3001	Implantable metal	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT3002	Implantable synthetic polymer	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT3006	Implantable inorganic material	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT314	Electrically powered	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT321	Long-term surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT449	Generators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT981	Single use	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT983	Surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
13019	2023		47265	Dual-chamber implantable pacemaker, rate-responsive	CT141	Electrodes/Leads and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	21/04/2023	21/04/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	21/04/2023	21/04/2023
					CT145	Cardiovascular devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	21/04/2023	21/04/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	21/04/2023	21/04/2023
					CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	21/04/2023	21/04/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	21/04/2023	21/04/2023
					CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	21/04/2023	21/04/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	21/04/2023	21/04/2023

					CT233	Surgical	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	21/04/2023	21/04/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	21/04/2023	21/04/2023
					CT2406	Implantable	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	21/04/2023	21/04/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	21/04/2023	21/04/2023
					CT2407	Active implantable	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	21/04/2023	21/04/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	21/04/2023	21/04/2023
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	21/04/2023	21/04/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	21/04/2023	21/04/2023
					CT2727	Active implant devices/ancillaries	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	21/04/2023	21/04/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	21/04/2023	21/04/2023
					CT2750	Implantable cardiac electrical stimulators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	21/04/2023	21/04/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	21/04/2023	21/04/2023
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	21/04/2023	21/04/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	21/04/2023	21/04/2023
					CT2985	Active	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	21/04/2023	21/04/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	21/04/2023	21/04/2023
					CT3001	Implantable metal	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	21/04/2023	21/04/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	21/04/2023	21/04/2023
					CT3002	Implantable synthetic polymer	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	21/04/2023	21/04/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	21/04/2023	21/04/2023
					CT3006	Implantable inorganic material	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	21/04/2023	21/04/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	21/04/2023	21/04/2023
					CT314	Electrically powered	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	21/04/2023	21/04/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	21/04/2023	21/04/2023
					CT321	Long-term surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	21/04/2023	21/04/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	21/04/2023	21/04/2023
					CT449	Generators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	21/04/2023	21/04/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	21/04/2023	21/04/2023
					CT981	Single use	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	21/04/2023	21/04/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	21/04/2023	21/04/2023
					CT983	Surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	21/04/2023	21/04/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	21/04/2023	21/04/2023
13020	2023	All other reportable incidents	47265	Dual-chamber implantable pacemaker, rate-responsive	CT141	Electrodes/Leads and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	21/04/2023	21/04/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	21/04/2023	21/04/2023
					CT145	Cardiovascular devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	21/04/2023	21/04/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	21/04/2023	21/04/2023
					CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	21/04/2023	21/04/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	21/04/2023	21/04/2023
					CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	21/04/2023	21/04/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	21/04/2023	21/04/2023
					CT233	Surgical	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	21/04/2023	21/04/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	21/04/2023	21/04/2023
					CT2406	Implantable	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	21/04/2023	21/04/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	21/04/2023	21/04/2023
					CT2407	Active implantable	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	21/04/2023	21/04/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	21/04/2023	21/04/2023
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	21/04/2023	21/04/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	21/04/2023	21/04/2023
					CT2727	Active implant devices/ancillaries	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	21/04/2023	21/04/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	21/04/2023	21/04/2023
					CT2750	Implantable cardiac electrical stimulators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	21/04/2023	21/04/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	21/04/2023	21/04/2023
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	21/04/2023	21/04/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	21/04/2023	21/04/2023
					CT2985	Active	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	21/04/2023	21/04/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	21/04/2023	21/04/2023
					CT3001	Implantable metal	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	21/04/2023	21/04/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	21/04/2023	21/04/2023
					CT3002	Implantable synthetic polymer	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	21/04/2023	21/04/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	21/04/2023	21/04/2023
					CT3006	Implantable inorganic material	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	21/04/2023	21/04/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	21/04/2023	21/04/2023
					CT314	Electrically powered	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	21/04/2023	21/04/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	21/04/2023	21/04/2023
					CT321	Long-term surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	21/04/2023	21/04/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	21/04/2023	21/04/2023
					CT449	Generators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	21/04/2023	21/04/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	21/04/2023	21/04/2023
					CT981	Single use	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	21/04/2023	21/04/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	21/04/2023	21/04/2023
					CT983	Surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	21/04/2023	21/04/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	21/04/2023	21/04/2023
13021	2023	All other reportable incidents	47265	Dual-chamber implantable pacemaker, rate-responsive	CT141	Electrodes/Leads and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A160601	Premature Elective Replacement Indicator		
							F08	Hospitalization or Prolonged Hospitalization	A160601	Premature Elective Replacement Indicator		
					CT145	Cardiovascular devices	E2403	No Clinical Signs, Symptoms or Conditions	A160601	Premature Elective Replacement Indicator		
							F08	Hospitalization or Prolonged Hospitalization	A160601	Premature Elective Replacement Indicator		
					CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A160601	Premature Elective Replacement Indicator		
							F08	Hospitalization or Prolonged Hospitalization	A160601	Premature Elective Replacement Indicator		
					CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A160601	Premature Elective Replacement Indicator		
							F08	Hospitalization or Prolonged Hospitalization	A160601	Premature Elective Replacement Indicator		
					CT233	Surgical	E2403	No Clinical Signs, Symptoms or Conditions	A160601	Premature Elective Replacement Indicator		
							F08	Hospitalization or Prolonged Hospitalization	A160601	Premature Elective Replacement Indicator		
					CT2406	Implantable	E2403	No Clinical Signs, Symptoms or Conditions	A160601	Premature Elective Replacement Indicator		
							F08	Hospitalization or Prolonged Hospitalization	A160601	Premature Elective Replacement Indicator		
					CT2407	Active implantable	E2403	No Clinical Signs, Symptoms or Conditions	A160601	Premature Elective Replacement Indicator		
							F08	Hospitalization or Prolonged Hospitalization	A160601	Premature Elective Replacement Indicator		
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A160601	Premature Elective Replacement Indicator		
							F08	Hospitalization or Prolonged Hospitalization	A160601	Premature Elective Replacement Indicator		

					CT2727	Active implant devices/ancillaries	E2403	No Clinical Signs, Symptoms or Conditions	A160601	Premature Elective Replacement Indicator		
							F08	Hospitalization or Prolonged Hospitalization	A160601	Premature Elective Replacement Indicator		
					CT2750	Implantable cardiac electrical stimulators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A160601	Premature Elective Replacement Indicator		
							F08	Hospitalization or Prolonged Hospitalization	A160601	Premature Elective Replacement Indicator		
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A160601	Premature Elective Replacement Indicator		
							F08	Hospitalization or Prolonged Hospitalization	A160601	Premature Elective Replacement Indicator		
					CT2985	Active	E2403	No Clinical Signs, Symptoms or Conditions	A160601	Premature Elective Replacement Indicator		
							F08	Hospitalization or Prolonged Hospitalization	A160601	Premature Elective Replacement Indicator		
					CT3001	Implantable metal	E2403	No Clinical Signs, Symptoms or Conditions	A160601	Premature Elective Replacement Indicator		
							F08	Hospitalization or Prolonged Hospitalization	A160601	Premature Elective Replacement Indicator		
					CT3002	Implantable synthetic polymer	E2403	No Clinical Signs, Symptoms or Conditions	A160601	Premature Elective Replacement Indicator		
							F08	Hospitalization or Prolonged Hospitalization	A160601	Premature Elective Replacement Indicator		
					CT3006	Implantable inorganic material	E2403	No Clinical Signs, Symptoms or Conditions	A160601	Premature Elective Replacement Indicator		
							F08	Hospitalization or Prolonged Hospitalization	A160601	Premature Elective Replacement Indicator		
					CT314	Electrically powered	E2403	No Clinical Signs, Symptoms or Conditions	A160601	Premature Elective Replacement Indicator		
							F08	Hospitalization or Prolonged Hospitalization	A160601	Premature Elective Replacement Indicator		
					CT321	Long-term surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A160601	Premature Elective Replacement Indicator		
							F08	Hospitalization or Prolonged Hospitalization	A160601	Premature Elective Replacement Indicator		
					CT449	Generators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A160601	Premature Elective Replacement Indicator		
							F08	Hospitalization or Prolonged Hospitalization	A160601	Premature Elective Replacement Indicator		
					CT981	Single use	E2403	No Clinical Signs, Symptoms or Conditions	A160601	Premature Elective Replacement Indicator		
							F08	Hospitalization or Prolonged Hospitalization	A160601	Premature Elective Replacement Indicator		
					CT983	Surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A160601	Premature Elective Replacement Indicator		
							F08	Hospitalization or Prolonged Hospitalization	A160601	Premature Elective Replacement Indicator		
13022	2023	All other reportable incidents	35223	Endocardial/interventricular septal pacing lead	CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070909	Oversensing		
							F26	No Health Consequences or Impact	A070909	Oversensing		
13023	2023	All other reportable incidents	17882	Physiologic-monitoring defibrillation system	CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A071301	Failure to Deliver Shock/Stimulation		
							F26	No Health Consequences or Impact	A071301	Failure to Deliver Shock/Stimulation		
13024	2023	All other reportable incidents	47270	Cardiac resynchronization therapy implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A070909	Oversensing		
									A071204	No Pacing		
							F23	Unexpected Medical Intervention	A070909	Oversensing		
									A071204	No Pacing		
13025	2023	All other reportable incidents	35223	Endocardial/interventricular septal pacing lead	CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070909	Oversensing		
							F1901	Additional Surgery	A070909	Oversensing		
13026	2023	All other reportable incidents	47265	Dual-chamber implantable pacemaker, rate-responsive	CT141	Electrodes/Leads and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A071102	Failure to Interrogate	06/02/2023	06/02/2023
									A071204	No Pacing	06/02/2023	06/02/2023
							F1901	Additional Surgery	A071102	Failure to Interrogate	06/02/2023	06/02/2023
									A071204	No Pacing	06/02/2023	06/02/2023
					CT145	Cardiovascular devices	E2403	No Clinical Signs, Symptoms or Conditions	A071102	Failure to Interrogate	06/02/2023	06/02/2023
									A071204	No Pacing	06/02/2023	06/02/2023
							F1901	Additional Surgery	A071102	Failure to Interrogate	06/02/2023	06/02/2023
									A071204	No Pacing	06/02/2023	06/02/2023
					CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A071102	Failure to Interrogate	06/02/2023	06/02/2023
									A071204	No Pacing	06/02/2023	06/02/2023
							F1901	Additional Surgery	A071102	Failure to Interrogate	06/02/2023	06/02/2023
									A071204	No Pacing	06/02/2023	06/02/2023
					CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A071102	Failure to Interrogate	06/02/2023	06/02/2023
									A071204	No Pacing	06/02/2023	06/02/2023
							F1901	Additional Surgery	A071102	Failure to Interrogate	06/02/2023	06/02/2023
									A071204	No Pacing	06/02/2023	06/02/2023
					CT233	Surgical	E2403	No Clinical Signs, Symptoms or Conditions	A071102	Failure to Interrogate	06/02/2023	06/02/2023
									A071204	No Pacing	06/02/2023	06/02/2023
							F1901	Additional Surgery	A071102	Failure to Interrogate	06/02/2023	06/02/2023
									A071204	No Pacing	06/02/2023	06/02/2023
					CT2406	Implantable	E2403	No Clinical Signs, Symptoms or Conditions	A071102	Failure to Interrogate	06/02/2023	06/02/2023
									A071204	No Pacing	06/02/2023	06/02/2023
							F1901	Additional Surgery	A071102	Failure to Interrogate	06/02/2023	06/02/2023
									A071204	No Pacing	06/02/2023	06/02/2023
					CT2407	Active implantable	E2403	No Clinical Signs, Symptoms or Conditions	A071102	Failure to Interrogate	06/02/2023	06/02/2023
									A071204	No Pacing	06/02/2023	06/02/2023
							F1901	Additional Surgery	A071102	Failure to Interrogate	06/02/2023	06/02/2023
									A071204	No Pacing	06/02/2023	06/02/2023
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A071102	Failure to Interrogate	06/02/2023	06/02/2023
									A071204	No Pacing	06/02/2023	06/02/2023
							F1901	Additional Surgery	A071102	Failure to Interrogate	06/02/2023	06/02/2023
									A071204	No Pacing	06/02/2023	06/02/2023
					CT2727	Active implant devices/ancillaries	E2403	No Clinical Signs, Symptoms or Conditions	A071102	Failure to Interrogate	06/02/2023	06/02/2023
									A071204	No Pacing	06/02/2023	06/02/2023
							F1901	Additional Surgery	A071102	Failure to Interrogate	06/02/2023	06/02/2023
									A071204	No Pacing	06/02/2023	06/02/2023
					CT2750	Implantable cardiac electrical stimulators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A071102	Failure to Interrogate	06/02/2023	06/02/2023
									A071204	No Pacing	06/02/2023	06/02/2023
							F1901	Additional Surgery	A071102	Failure to Interrogate	06/02/2023	06/02/2023
									A071204	No Pacing	06/02/2023	06/02/2023
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A071102	Failure to Interrogate	06/02/2023	06/02/2023
									A071204	No Pacing	06/02/2023	06/02/2023
							F1901	Additional Surgery	A071102	Failure to Interrogate	06/02/2023	06/02/2023
									A071204	No Pacing	06/02/2023	06/02/2023
					CT2985	Active	E2403	No Clinical Signs, Symptoms or Conditions	A071102	Failure to Interrogate	06/02/2023	06/02/2023
									A071204	No Pacing	06/02/2023	06/02/2023
							F1901	Additional Surgery	A071102	Failure to Interrogate	06/02/2023	06/02/2023
									A071204	No Pacing	06/02/2023	06/02/2023
					CT3001	Implantable metal	E2403	No Clinical Signs, Symptoms or Conditions	A071102	Failure to Interrogate	06/02/2023	06/02/2023
									A071204	No Pacing	06/02/2023	06/02/2023
							F1901	Additional Surgery	A071102	Failure to Interrogate	06/02/2023	06/02/2023
									A071204	No Pacing	06/02/2023	06/02/2023
					CT3002	Implantable synthetic polymer	E2403	No Clinical Signs, Symptoms or Conditions	A071102	Failure to Interrogate	06/02/2023	06/02/2023
									A071204	No Pacing	06/02/2023	06/02/2023

							F1901	Additional Surgery	A071102	Failure to Interrogate	06/02/2023	06/02/2023
									A071204	No Pacing	06/02/2023	06/02/2023
					CT3006	Implantable inorganic material	E2403	No Clinical Signs, Symptoms or Conditions	A071102	Failure to Interrogate	06/02/2023	06/02/2023
									A071204	No Pacing	06/02/2023	06/02/2023
							F1901	Additional Surgery	A071102	Failure to Interrogate	06/02/2023	06/02/2023
									A071204	No Pacing	06/02/2023	06/02/2023
					CT314	Electrically powered	E2403	No Clinical Signs, Symptoms or Conditions	A071102	Failure to Interrogate	06/02/2023	06/02/2023
									A071204	No Pacing	06/02/2023	06/02/2023
							F1901	Additional Surgery	A071102	Failure to Interrogate	06/02/2023	06/02/2023
									A071204	No Pacing	06/02/2023	06/02/2023
					CT321	Long-term surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A071102	Failure to Interrogate	06/02/2023	06/02/2023
									A071204	No Pacing	06/02/2023	06/02/2023
							F1901	Additional Surgery	A071102	Failure to Interrogate	06/02/2023	06/02/2023
									A071204	No Pacing	06/02/2023	06/02/2023
					CT449	Generators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A071102	Failure to Interrogate	06/02/2023	06/02/2023
									A071204	No Pacing	06/02/2023	06/02/2023
							F1901	Additional Surgery	A071102	Failure to Interrogate	06/02/2023	06/02/2023
									A071204	No Pacing	06/02/2023	06/02/2023
					CT981	Single use	E2403	No Clinical Signs, Symptoms or Conditions	A071102	Failure to Interrogate	06/02/2023	06/02/2023
									A071204	No Pacing	06/02/2023	06/02/2023
							F1901	Additional Surgery	A071102	Failure to Interrogate	06/02/2023	06/02/2023
									A071204	No Pacing	06/02/2023	06/02/2023
					CT983	Surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A071102	Failure to Interrogate	06/02/2023	06/02/2023
									A071204	No Pacing	06/02/2023	06/02/2023
							F1901	Additional Surgery	A071102	Failure to Interrogate	06/02/2023	06/02/2023
									A071204	No Pacing	06/02/2023	06/02/2023
13027	2023	All other reportable incidents	17882	Physiologic-monitoring defibrillation system	CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A150105	Key or Button Unresponsive/not Working		
							F27	Problem Identified During Non-Clinical Procedure	A150105	Key or Button Unresponsive/not Working		
13028	2023	All other reportable incidents	35223	Endocardial/interventricular septal pacing lead	CT1684	Cardiac pacemakers and associated devices	E011902	Presyncope	A070101	Failure to Capture	01/02/2023	01/02/2023
									A070102	High Capture Threshold	01/02/2023	01/02/2023
									A070104	Unstable Capture Threshold	01/02/2023	01/02/2023
							E011903	Syncope/Fainting	A070101	Failure to Capture	01/02/2023	01/02/2023
									A070102	High Capture Threshold	01/02/2023	01/02/2023
									A070104	Unstable Capture Threshold	01/02/2023	01/02/2023
							F08	Hospitalization or Prolonged Hospitalization	A070101	Failure to Capture	01/02/2023	01/02/2023
									A070102	High Capture Threshold	01/02/2023	01/02/2023
									A070104	Unstable Capture Threshold	01/02/2023	01/02/2023
							F1203	Life Threatening Illness or Injury	A070101	Failure to Capture	01/02/2023	01/02/2023
									A070102	High Capture Threshold	01/02/2023	01/02/2023
									A070104	Unstable Capture Threshold	01/02/2023	01/02/2023
							F1905	Device Revision or Replacement	A070101	Failure to Capture	01/02/2023	01/02/2023
									A070102	High Capture Threshold	01/02/2023	01/02/2023
									A070104	Unstable Capture Threshold	01/02/2023	01/02/2023
13029	2023	All other reportable incidents	35223	Endocardial/interventricular septal pacing lead	CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A27	Appropriate Device Problem Term/Code not Available	01/02/2023	01/02/2023
							F12	Serious Injury/ Illness/ Impairment	A27	Appropriate Device Problem Term/Code not Available	01/02/2023	01/02/2023
							F1905	Device Revision or Replacement	A27	Appropriate Device Problem Term/Code not Available	01/02/2023	01/02/2023
13030	2023	All other reportable incidents	35223	Endocardial/interventricular septal pacing lead	CT1684	Cardiac pacemakers and associated devices	E0306	Sepsis	A0722	Impedance Problem	01/02/2023	28/02/2023
									A072201	High Impedance	01/02/2023	28/02/2023
							E1718	Skin Erosion	A0722	Impedance Problem	01/02/2023	28/02/2023
									A072201	High Impedance	01/02/2023	28/02/2023
							E2403	No Clinical Signs, Symptoms or Conditions	A0722	Impedance Problem	01/02/2023	28/02/2023
									A072201	High Impedance	01/02/2023	28/02/2023
							F08	Hospitalization or Prolonged Hospitalization	A0722	Impedance Problem	01/02/2023	28/02/2023
									A072201	High Impedance	01/02/2023	28/02/2023
							F1203	Life Threatening Illness or Injury	A0722	Impedance Problem	01/02/2023	28/02/2023
									A072201	High Impedance	01/02/2023	28/02/2023
							F1903	Device Explantation	A0722	Impedance Problem	01/02/2023	28/02/2023
									A072201	High Impedance	01/02/2023	28/02/2023
							F26	No Health Consequences or Impact	A0722	Impedance Problem	01/02/2023	28/02/2023
									A072201	High Impedance	01/02/2023	28/02/2023
13031	2023	All other reportable incidents	35223	Endocardial/interventricular septal pacing lead	CT1684	Cardiac pacemakers and associated devices	E0306	Sepsis	A0722	Impedance Problem	01/02/2023	28/02/2023
									A072201	High Impedance	01/02/2023	28/02/2023
							E1718	Skin Erosion	A0722	Impedance Problem	01/02/2023	28/02/2023
									A072201	High Impedance	01/02/2023	28/02/2023
							E2403	No Clinical Signs, Symptoms or Conditions	A0722	Impedance Problem	01/02/2023	28/02/2023
									A072201	High Impedance	01/02/2023	28/02/2023
							F08	Hospitalization or Prolonged Hospitalization	A0722	Impedance Problem	01/02/2023	28/02/2023
									A072201	High Impedance	01/02/2023	28/02/2023
							F1203	Life Threatening Illness or Injury	A0722	Impedance Problem	01/02/2023	28/02/2023
									A072201	High Impedance	01/02/2023	28/02/2023
							F1903	Device Explantation	A0722	Impedance Problem	01/02/2023	28/02/2023
									A072201	High Impedance	01/02/2023	28/02/2023
							F26	No Health Consequences or Impact	A0722	Impedance Problem	01/02/2023	28/02/2023
									A072201	High Impedance	01/02/2023	28/02/2023
13032	2023	All other reportable incidents	35852	Single-chamber implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
13033	2023	All other reportable incidents	47205	Cardiac pulse generator programmer	CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A150105	Key or Button Unresponsive/not Working		
							F26	No Health Consequences or Impact	A150105	Key or Button Unresponsive/not Working		
13034	2023	All other reportable incidents	36241	Pacemaker/defibrillator lead	CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A040101	Fracture		
									A070101	Failure to Capture		
									A072201	High Impedance		
							F2203	Imaging Required	A040101	Fracture		
									A070101	Failure to Capture		
									A072201	High Impedance		
							F26	No Health Consequences or Impact	A040101	Fracture		

									A070101	Failure to Capture		
									A072201	High Impedance		
13035	2023	All other reportable incidents	36241	Pacemaker/defibrillator lead	CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A0203	Defective Device		
									A070909	Oversensing		
									A090801	Signal Artifact/Noise		
							F08	Hospitalization or Prolonged Hospitalization	A0203	Defective Device		
									A070909	Oversensing		
									A090801	Signal Artifact/Noise		
							F19	Surgical Intervention	A0203	Defective Device		
									A070909	Oversensing		
									A090801	Signal Artifact/Noise		
13036	2023	All other reportable incidents	35852	Single-chamber implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	09/03/2023	09/03/2023
									A0908	Incorrect, Inadequate or Imprecise Result or Readings	09/03/2023	09/03/2023
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	09/03/2023	09/03/2023
									A0908	Incorrect, Inadequate or Imprecise Result or Readings	09/03/2023	09/03/2023
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	09/03/2023	09/03/2023
									A0908	Incorrect, Inadequate or Imprecise Result or Readings	09/03/2023	09/03/2023
13037	2023	All other reportable incidents	35852	Single-chamber implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
13038	2023	All other reportable incidents	47265	Dual-chamber implantable pacemaker, rate-responsive	CT141	Electrodes/Leads and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem	26/04/2023	26/04/2023
									A070504	Premature Discharge of Battery	26/04/2023	26/04/2023
							F08	Hospitalization or Prolonged Hospitalization	A0705	Battery Problem	26/04/2023	26/04/2023
									A070504	Premature Discharge of Battery	26/04/2023	26/04/2023
							F1905	Device Revision or Replacement	A0705	Battery Problem	26/04/2023	26/04/2023
									A070504	Premature Discharge of Battery	26/04/2023	26/04/2023
					CT145	Cardiovascular devices	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem	26/04/2023	26/04/2023
									A070504	Premature Discharge of Battery	26/04/2023	26/04/2023
							F08	Hospitalization or Prolonged Hospitalization	A0705	Battery Problem	26/04/2023	26/04/2023
									A070504	Premature Discharge of Battery	26/04/2023	26/04/2023
							F1905	Device Revision or Replacement	A0705	Battery Problem	26/04/2023	26/04/2023
									A070504	Premature Discharge of Battery	26/04/2023	26/04/2023
					CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem	26/04/2023	26/04/2023
									A070504	Premature Discharge of Battery	26/04/2023	26/04/2023
							F08	Hospitalization or Prolonged Hospitalization	A0705	Battery Problem	26/04/2023	26/04/2023
									A070504	Premature Discharge of Battery	26/04/2023	26/04/2023
							F1905	Device Revision or Replacement	A0705	Battery Problem	26/04/2023	26/04/2023
									A070504	Premature Discharge of Battery	26/04/2023	26/04/2023
					CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem	26/04/2023	26/04/2023
									A070504	Premature Discharge of Battery	26/04/2023	26/04/2023
							F08	Hospitalization or Prolonged Hospitalization	A0705	Battery Problem	26/04/2023	26/04/2023
									A070504	Premature Discharge of Battery	26/04/2023	26/04/2023
							F1905	Device Revision or Replacement	A0705	Battery Problem	26/04/2023	26/04/2023
									A070504	Premature Discharge of Battery	26/04/2023	26/04/2023
					CT233	Surgical	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem	26/04/2023	26/04/2023
									A070504	Premature Discharge of Battery	26/04/2023	26/04/2023
							F08	Hospitalization or Prolonged Hospitalization	A0705	Battery Problem	26/04/2023	26/04/2023
									A070504	Premature Discharge of Battery	26/04/2023	26/04/2023
							F1905	Device Revision or Replacement	A0705	Battery Problem	26/04/2023	26/04/2023
									A070504	Premature Discharge of Battery	26/04/2023	26/04/2023
					CT2406	Implantable	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem	26/04/2023	26/04/2023
									A070504	Premature Discharge of Battery	26/04/2023	26/04/2023
							F08	Hospitalization or Prolonged Hospitalization	A0705	Battery Problem	26/04/2023	26/04/2023
									A070504	Premature Discharge of Battery	26/04/2023	26/04/2023
							F1905	Device Revision or Replacement	A0705	Battery Problem	26/04/2023	26/04/2023
									A070504	Premature Discharge of Battery	26/04/2023	26/04/2023
					CT2407	Active implantable	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem	26/04/2023	26/04/2023
									A070504	Premature Discharge of Battery	26/04/2023	26/04/2023
							F08	Hospitalization or Prolonged Hospitalization	A0705	Battery Problem	26/04/2023	26/04/2023
									A070504	Premature Discharge of Battery	26/04/2023	26/04/2023
							F1905	Device Revision or Replacement	A0705	Battery Problem	26/04/2023	26/04/2023
									A070504	Premature Discharge of Battery	26/04/2023	26/04/2023
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem	26/04/2023	26/04/2023
									A070504	Premature Discharge of Battery	26/04/2023	26/04/2023
							F08	Hospitalization or Prolonged Hospitalization	A0705	Battery Problem	26/04/2023	26/04/2023
									A070504	Premature Discharge of Battery	26/04/2023	26/04/2023
							F1905	Device Revision or Replacement	A0705	Battery Problem	26/04/2023	26/04/2023
									A070504	Premature Discharge of Battery	26/04/2023	26/04/2023
					CT2727	Active implant devices/ancillaries	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem	26/04/2023	26/04/2023
									A070504	Premature Discharge of Battery	26/04/2023	26/04/2023
							F08	Hospitalization or Prolonged Hospitalization	A0705	Battery Problem	26/04/2023	26/04/2023
									A070504	Premature Discharge of Battery	26/04/2023	26/04/2023
							F1905	Device Revision or Replacement	A0705	Battery Problem	26/04/2023	26/04/2023
									A070504	Premature Discharge of Battery	26/04/2023	26/04/2023
					CT2750	Implantable cardiac electrical stimulators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem	26/04/2023	26/04/2023
									A070504	Premature Discharge of Battery	26/04/2023	26/04/2023
							F08	Hospitalization or Prolonged Hospitalization	A0705	Battery Problem	26/04/2023	26/04/2023
									A070504	Premature Discharge of Battery	26/04/2023	26/04/2023
							F1905	Device Revision or Replacement	A0705	Battery Problem	26/04/2023	26/04/2023
									A070504	Premature Discharge of Battery	26/04/2023	26/04/2023
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem	26/04/2023	26/04/2023
									A070504	Premature Discharge of Battery	26/04/2023	26/04/2023
							F08	Hospitalization or Prolonged Hospitalization	A0705	Battery Problem	26/04/2023	26/04/2023
									A070504	Premature Discharge of Battery	26/04/2023	26/04/2023
							F1905	Device Revision or Replacement	A0705	Battery Problem	26/04/2023	26/04/2023
									A070504	Premature Discharge of Battery	26/04/2023	26/04/2023

					CT2985	Active	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem	26/04/2023	26/04/2023
									A070504	Premature Discharge of Battery	26/04/2023	26/04/2023
							F08	Hospitalization or Prolonged Hospitalization	A0705	Battery Problem	26/04/2023	26/04/2023
									A070504	Premature Discharge of Battery	26/04/2023	26/04/2023
							F1905	Device Revision or Replacement	A0705	Battery Problem	26/04/2023	26/04/2023
									A070504	Premature Discharge of Battery	26/04/2023	26/04/2023
					CT3001	Implantable metal	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem	26/04/2023	26/04/2023
									A070504	Premature Discharge of Battery	26/04/2023	26/04/2023
							F08	Hospitalization or Prolonged Hospitalization	A0705	Battery Problem	26/04/2023	26/04/2023
									A070504	Premature Discharge of Battery	26/04/2023	26/04/2023
							F1905	Device Revision or Replacement	A0705	Battery Problem	26/04/2023	26/04/2023
									A070504	Premature Discharge of Battery	26/04/2023	26/04/2023
					CT3002	Implantable synthetic polymer	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem	26/04/2023	26/04/2023
									A070504	Premature Discharge of Battery	26/04/2023	26/04/2023
							F08	Hospitalization or Prolonged Hospitalization	A0705	Battery Problem	26/04/2023	26/04/2023
									A070504	Premature Discharge of Battery	26/04/2023	26/04/2023
							F1905	Device Revision or Replacement	A0705	Battery Problem	26/04/2023	26/04/2023
									A070504	Premature Discharge of Battery	26/04/2023	26/04/2023
					CT3006	Implantable inorganic material	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem	26/04/2023	26/04/2023
									A070504	Premature Discharge of Battery	26/04/2023	26/04/2023
							F08	Hospitalization or Prolonged Hospitalization	A0705	Battery Problem	26/04/2023	26/04/2023
									A070504	Premature Discharge of Battery	26/04/2023	26/04/2023
							F1905	Device Revision or Replacement	A0705	Battery Problem	26/04/2023	26/04/2023
									A070504	Premature Discharge of Battery	26/04/2023	26/04/2023
					CT314	Electrically powered	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem	26/04/2023	26/04/2023
									A070504	Premature Discharge of Battery	26/04/2023	26/04/2023
							F08	Hospitalization or Prolonged Hospitalization	A0705	Battery Problem	26/04/2023	26/04/2023
									A070504	Premature Discharge of Battery	26/04/2023	26/04/2023
							F1905	Device Revision or Replacement	A0705	Battery Problem	26/04/2023	26/04/2023
									A070504	Premature Discharge of Battery	26/04/2023	26/04/2023
					CT321	Long-term surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem	26/04/2023	26/04/2023
									A070504	Premature Discharge of Battery	26/04/2023	26/04/2023
							F08	Hospitalization or Prolonged Hospitalization	A0705	Battery Problem	26/04/2023	26/04/2023
									A070504	Premature Discharge of Battery	26/04/2023	26/04/2023
							F1905	Device Revision or Replacement	A0705	Battery Problem	26/04/2023	26/04/2023
									A070504	Premature Discharge of Battery	26/04/2023	26/04/2023
					CT449	Generators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem	26/04/2023	26/04/2023
									A070504	Premature Discharge of Battery	26/04/2023	26/04/2023
							F08	Hospitalization or Prolonged Hospitalization	A0705	Battery Problem	26/04/2023	26/04/2023
									A070504	Premature Discharge of Battery	26/04/2023	26/04/2023
							F1905	Device Revision or Replacement	A0705	Battery Problem	26/04/2023	26/04/2023
									A070504	Premature Discharge of Battery	26/04/2023	26/04/2023
					CT981	Single use	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem	26/04/2023	26/04/2023
									A070504	Premature Discharge of Battery	26/04/2023	26/04/2023
							F08	Hospitalization or Prolonged Hospitalization	A0705	Battery Problem	26/04/2023	26/04/2023
									A070504	Premature Discharge of Battery	26/04/2023	26/04/2023
							F1905	Device Revision or Replacement	A0705	Battery Problem	26/04/2023	26/04/2023
									A070504	Premature Discharge of Battery	26/04/2023	26/04/2023
					CT983	Surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem	26/04/2023	26/04/2023
									A070504	Premature Discharge of Battery	26/04/2023	26/04/2023
							F08	Hospitalization or Prolonged Hospitalization	A0705	Battery Problem	26/04/2023	26/04/2023
									A070504	Premature Discharge of Battery	26/04/2023	26/04/2023
							F1905	Device Revision or Replacement	A0705	Battery Problem	26/04/2023	26/04/2023
									A070504	Premature Discharge of Battery	26/04/2023	26/04/2023
13039	2023	All other reportable incidents	37265	Dual-chamber implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A0713	Defibrillation/Stimulation Problem		
									A0722	Impedance Problem		
									A090804	False Positive Result		
									A090805	Incorrect Measurement		
							F08	Hospitalization or Prolonged Hospitalization	A0713	Defibrillation/Stimulation Problem		
									A0722	Impedance Problem		
									A090804	False Positive Result		
									A090805	Incorrect Measurement		
							F12	Serious Injury/ Illness/ Impairment	A0713	Defibrillation/Stimulation Problem		
									A0722	Impedance Problem		
									A090804	False Positive Result		
									A090805	Incorrect Measurement		
							F1905	Device Revision or Replacement	A0713	Defibrillation/Stimulation Problem		
									A0722	Impedance Problem		
									A090804	False Positive Result		
									A090805	Incorrect Measurement		
13040	2023	All other reportable incidents	47265	Dual-chamber implantable pacemaker, rate-responsive	CT141	Electrodes/Leads and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	16/03/2023	16/03/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	16/03/2023	16/03/2023
					CT145	Cardiovascular devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	16/03/2023	16/03/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	16/03/2023	16/03/2023
					CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	16/03/2023	16/03/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	16/03/2023	16/03/2023
					CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	16/03/2023	16/03/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	16/03/2023	16/03/2023
					CT233	Surgical	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	16/03/2023	16/03/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	16/03/2023	16/03/2023
					CT2406	Implantable	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	16/03/2023	16/03/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	16/03/2023	16/03/2023
					CT2407	Active implantable	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	16/03/2023	16/03/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	16/03/2023	16/03/2023
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	16/03/2023	16/03/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	16/03/2023	16/03/2023
					CT2727	Active implant devices/ancillaries	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	16/03/2023	16/03/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	16/03/2023	16/03/2023

					CT2750	Implantable cardiac electrical stimulators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	16/03/2023	16/03/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	16/03/2023	16/03/2023
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	16/03/2023	16/03/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	16/03/2023	16/03/2023
					CT2985	Active	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	16/03/2023	16/03/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	16/03/2023	16/03/2023
					CT3001	Implantable metal	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	16/03/2023	16/03/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	16/03/2023	16/03/2023
					CT3002	Implantable synthetic polymer	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	16/03/2023	16/03/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	16/03/2023	16/03/2023
					CT3006	Implantable inorganic material	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	16/03/2023	16/03/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	16/03/2023	16/03/2023
					CT314	Electrically powered	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	16/03/2023	16/03/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	16/03/2023	16/03/2023
					CT321	Long-term surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	16/03/2023	16/03/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	16/03/2023	16/03/2023
					CT449	Generators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	16/03/2023	16/03/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	16/03/2023	16/03/2023
					CT981	Single use	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	16/03/2023	16/03/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	16/03/2023	16/03/2023
					CT983	Surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	16/03/2023	16/03/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	16/03/2023	16/03/2023
13041	2023	All other reportable incidents	47265	Dual-chamber implantable pacemaker, rate-responsive	CT141	Electrodes/Leads and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	17/03/2023	17/03/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	17/03/2023	17/03/2023
					CT145	Cardiovascular devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	17/03/2023	17/03/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	17/03/2023	17/03/2023
					CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	17/03/2023	17/03/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	17/03/2023	17/03/2023
					CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	17/03/2023	17/03/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	17/03/2023	17/03/2023
					CT233	Surgical	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	17/03/2023	17/03/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	17/03/2023	17/03/2023
					CT2406	Implantable	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	17/03/2023	17/03/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	17/03/2023	17/03/2023
					CT2407	Active implantable	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	17/03/2023	17/03/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	17/03/2023	17/03/2023
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	17/03/2023	17/03/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	17/03/2023	17/03/2023
					CT2727	Active implant devices/ancillaries	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	17/03/2023	17/03/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	17/03/2023	17/03/2023
					CT2750	Implantable cardiac electrical stimulators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	17/03/2023	17/03/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	17/03/2023	17/03/2023
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	17/03/2023	17/03/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	17/03/2023	17/03/2023
					CT2985	Active	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	17/03/2023	17/03/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	17/03/2023	17/03/2023
					CT3001	Implantable metal	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	17/03/2023	17/03/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	17/03/2023	17/03/2023
					CT3002	Implantable synthetic polymer	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	17/03/2023	17/03/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	17/03/2023	17/03/2023
					CT3006	Implantable inorganic material	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	17/03/2023	17/03/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	17/03/2023	17/03/2023
					CT314	Electrically powered	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	17/03/2023	17/03/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	17/03/2023	17/03/2023
					CT321	Long-term surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	17/03/2023	17/03/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	17/03/2023	17/03/2023
					CT449	Generators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	17/03/2023	17/03/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	17/03/2023	17/03/2023
					CT981	Single use	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	17/03/2023	17/03/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	17/03/2023	17/03/2023
					CT983	Surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	17/03/2023	17/03/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	17/03/2023	17/03/2023
13042	2023		47265	Dual-chamber implantable pacemaker, rate-responsive	CT141	Electrodes/Leads and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	03/04/2023	03/04/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	03/04/2023	03/04/2023
					CT145	Cardiovascular devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	03/04/2023	03/04/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	03/04/2023	03/04/2023
					CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	03/04/2023	03/04/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	03/04/2023	03/04/2023
					CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	03/04/2023	03/04/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	03/04/2023	03/04/2023
					CT233	Surgical	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	03/04/2023	03/04/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	03/04/2023	03/04/2023
					CT2406	Implantable	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	03/04/2023	03/04/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	03/04/2023	03/04/2023
					CT2407	Active implantable	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	03/04/2023	03/04/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	03/04/2023	03/04/2023
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	03/04/2023	03/04/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	03/04/2023	03/04/2023
					CT2727	Active implant devices/ancillaries	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	03/04/2023	03/04/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	03/04/2023	03/04/2023
					CT2750	Implantable cardiac electrical stimulators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	03/04/2023	03/04/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	03/04/2023	03/04/2023
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	03/04/2023	03/04/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	03/04/2023	03/04/2023
					CT2985	Active	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	03/04/2023	03/04/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	03/04/2023	03/04/2023
					CT3001	Implantable metal	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	03/04/2023	03/04/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	03/04/2023	03/04/2023

					CT3002	Implantable synthetic polymer	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	03/04/2023	03/04/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	03/04/2023	03/04/2023
					CT3006	Implantable inorganic material	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	03/04/2023	03/04/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	03/04/2023	03/04/2023
					CT314	Electrically powered	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	03/04/2023	03/04/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	03/04/2023	03/04/2023
					CT321	Long-term surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	03/04/2023	03/04/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	03/04/2023	03/04/2023
					CT449	Generators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	03/04/2023	03/04/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	03/04/2023	03/04/2023
					CT981	Single use	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	03/04/2023	03/04/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	03/04/2023	03/04/2023
					CT983	Surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	03/04/2023	03/04/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	03/04/2023	03/04/2023
13043	2023		47265	Dual-chamber implantable pacemaker, rate-responsive	CT141	Electrodes/Leads and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	13/04/2023	13/04/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	13/04/2023	13/04/2023
					CT145	Cardiovascular devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	13/04/2023	13/04/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	13/04/2023	13/04/2023
					CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	13/04/2023	13/04/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	13/04/2023	13/04/2023
					CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	13/04/2023	13/04/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	13/04/2023	13/04/2023
					CT233	Surgical	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	13/04/2023	13/04/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	13/04/2023	13/04/2023
					CT2406	Implantable	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	13/04/2023	13/04/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	13/04/2023	13/04/2023
					CT2407	Active implantable	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	13/04/2023	13/04/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	13/04/2023	13/04/2023
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	13/04/2023	13/04/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	13/04/2023	13/04/2023
					CT2727	Active implant devices/ancillaries	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	13/04/2023	13/04/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	13/04/2023	13/04/2023
					CT2750	Implantable cardiac electrical stimulators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	13/04/2023	13/04/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	13/04/2023	13/04/2023
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	13/04/2023	13/04/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	13/04/2023	13/04/2023
					CT2985	Active	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	13/04/2023	13/04/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	13/04/2023	13/04/2023
					CT3001	Implantable metal	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	13/04/2023	13/04/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	13/04/2023	13/04/2023
					CT3002	Implantable synthetic polymer	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	13/04/2023	13/04/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	13/04/2023	13/04/2023
					CT3006	Implantable inorganic material	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	13/04/2023	13/04/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	13/04/2023	13/04/2023
					CT314	Electrically powered	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	13/04/2023	13/04/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	13/04/2023	13/04/2023
					CT321	Long-term surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	13/04/2023	13/04/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	13/04/2023	13/04/2023
					CT449	Generators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	13/04/2023	13/04/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	13/04/2023	13/04/2023
					CT981	Single use	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	13/04/2023	13/04/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	13/04/2023	13/04/2023
					CT983	Surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	13/04/2023	13/04/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	13/04/2023	13/04/2023
13044	2023		47265	Dual-chamber implantable pacemaker, rate-responsive	CT141	Electrodes/Leads and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	26/04/2023	26/04/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	26/04/2023	26/04/2023
					CT145	Cardiovascular devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	26/04/2023	26/04/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	26/04/2023	26/04/2023
					CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	26/04/2023	26/04/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	26/04/2023	26/04/2023
					CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	26/04/2023	26/04/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	26/04/2023	26/04/2023
					CT233	Surgical	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	26/04/2023	26/04/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	26/04/2023	26/04/2023
					CT2406	Implantable	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	26/04/2023	26/04/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	26/04/2023	26/04/2023
					CT2407	Active implantable	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	26/04/2023	26/04/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	26/04/2023	26/04/2023
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	26/04/2023	26/04/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	26/04/2023	26/04/2023
					CT2727	Active implant devices/ancillaries	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	26/04/2023	26/04/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	26/04/2023	26/04/2023
					CT2750	Implantable cardiac electrical stimulators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	26/04/2023	26/04/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	26/04/2023	26/04/2023
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	26/04/2023	26/04/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	26/04/2023	26/04/2023
					CT2985	Active	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	26/04/2023	26/04/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	26/04/2023	26/04/2023
					CT3001	Implantable metal	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	26/04/2023	26/04/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	26/04/2023	26/04/2023
					CT3002	Implantable synthetic polymer	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	26/04/2023	26/04/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	26/04/2023	26/04/2023
					CT3006	Implantable inorganic material	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	26/04/2023	26/04/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	26/04/2023	26/04/2023
					CT314	Electrically powered	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	26/04/2023	26/04/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	26/04/2023	26/04/2023
					CT321	Long-term surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	26/04/2023	26/04/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	26/04/2023	26/04/2023

					CT449	Generators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	26/04/2023	26/04/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	26/04/2023	26/04/2023
					CT981	Single use	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	26/04/2023	26/04/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	26/04/2023	26/04/2023
					CT983	Surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	26/04/2023	26/04/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	26/04/2023	26/04/2023
13045	2023	All other reportable incidents	47265	Dual-chamber implantable pacemaker, rate-responsive	CT141	Electrodes/Leads and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	08/03/2023	08/03/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	08/03/2023	08/03/2023
					CT145	Cardiovascular devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	08/03/2023	08/03/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	08/03/2023	08/03/2023
					CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	08/03/2023	08/03/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	08/03/2023	08/03/2023
					CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	08/03/2023	08/03/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	08/03/2023	08/03/2023
					CT233	Surgical	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	08/03/2023	08/03/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	08/03/2023	08/03/2023
					CT2406	Implantable	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	08/03/2023	08/03/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	08/03/2023	08/03/2023
					CT2407	Active implantable	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	08/03/2023	08/03/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	08/03/2023	08/03/2023
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	08/03/2023	08/03/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	08/03/2023	08/03/2023
					CT2727	Active implant devices/ancillaries	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	08/03/2023	08/03/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	08/03/2023	08/03/2023
					CT2750	Implantable cardiac electrical stimulators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	08/03/2023	08/03/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	08/03/2023	08/03/2023
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	08/03/2023	08/03/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	08/03/2023	08/03/2023
					CT2985	Active	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	08/03/2023	08/03/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	08/03/2023	08/03/2023
					CT3001	Implantable metal	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	08/03/2023	08/03/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	08/03/2023	08/03/2023
					CT3002	Implantable synthetic polymer	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	08/03/2023	08/03/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	08/03/2023	08/03/2023
					CT3006	Implantable inorganic material	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	08/03/2023	08/03/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	08/03/2023	08/03/2023
					CT314	Electrically powered	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	08/03/2023	08/03/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	08/03/2023	08/03/2023
					CT321	Long-term surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	08/03/2023	08/03/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	08/03/2023	08/03/2023
					CT449	Generators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	08/03/2023	08/03/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	08/03/2023	08/03/2023
					CT981	Single use	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	08/03/2023	08/03/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	08/03/2023	08/03/2023
					CT983	Surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	08/03/2023	08/03/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	08/03/2023	08/03/2023
13046	2023	All other reportable incidents	47265	Dual-chamber implantable pacemaker, rate-responsive	CT141	Electrodes/Leads and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	03/03/2023	03/03/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	03/03/2023	03/03/2023
					CT145	Cardiovascular devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	03/03/2023	03/03/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	03/03/2023	03/03/2023
					CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	03/03/2023	03/03/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	03/03/2023	03/03/2023
					CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	03/03/2023	03/03/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	03/03/2023	03/03/2023
					CT233	Surgical	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	03/03/2023	03/03/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	03/03/2023	03/03/2023
					CT2406	Implantable	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	03/03/2023	03/03/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	03/03/2023	03/03/2023
					CT2407	Active implantable	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	03/03/2023	03/03/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	03/03/2023	03/03/2023
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	03/03/2023	03/03/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	03/03/2023	03/03/2023
					CT2727	Active implant devices/ancillaries	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	03/03/2023	03/03/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	03/03/2023	03/03/2023
					CT2750	Implantable cardiac electrical stimulators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	03/03/2023	03/03/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	03/03/2023	03/03/2023
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	03/03/2023	03/03/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	03/03/2023	03/03/2023
					CT2985	Active	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	03/03/2023	03/03/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	03/03/2023	03/03/2023
					CT3001	Implantable metal	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	03/03/2023	03/03/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	03/03/2023	03/03/2023
					CT3002	Implantable synthetic polymer	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	03/03/2023	03/03/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	03/03/2023	03/03/2023
					CT3006	Implantable inorganic material	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	03/03/2023	03/03/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	03/03/2023	03/03/2023
					CT314	Electrically powered	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	03/03/2023	03/03/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	03/03/2023	03/03/2023
					CT321	Long-term surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	03/03/2023	03/03/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	03/03/2023	03/03/2023
					CT449	Generators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	03/03/2023	03/03/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	03/03/2023	03/03/2023
					CT981	Single use	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	03/03/2023	03/03/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	03/03/2023	03/03/2023
					CT983	Surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	03/03/2023	03/03/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	03/03/2023	03/03/2023
13047	2023		47265	Dual-chamber implantable pacemaker, rate-responsive	CT141	Electrodes/Leads and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem		
							F1901	Additional Surgery	A24	No Apparent Device Problem		

					CT145	Cardiovascular devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem		
							F1901	Additional Surgery	A24	No Apparent Device Problem		
					CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem		
							F1901	Additional Surgery	A24	No Apparent Device Problem		
					CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem		
							F1901	Additional Surgery	A24	No Apparent Device Problem		
					CT233	Surgical	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem		
							F1901	Additional Surgery	A24	No Apparent Device Problem		
					CT2406	Implantable	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem		
							F1901	Additional Surgery	A24	No Apparent Device Problem		
					CT2407	Active implantable	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem		
							F1901	Additional Surgery	A24	No Apparent Device Problem		
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem		
							F1901	Additional Surgery	A24	No Apparent Device Problem		
					CT2727	Active implant devices/ancillaries	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem		
							F1901	Additional Surgery	A24	No Apparent Device Problem		
					CT2750	Implantable cardiac electrical stimulators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem		
							F1901	Additional Surgery	A24	No Apparent Device Problem		
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem		
							F1901	Additional Surgery	A24	No Apparent Device Problem		
					CT2985	Active	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem		
							F1901	Additional Surgery	A24	No Apparent Device Problem		
					CT3001	Implantable metal	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem		
							F1901	Additional Surgery	A24	No Apparent Device Problem		
					CT3002	Implantable synthetic polymer	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem		
							F1901	Additional Surgery	A24	No Apparent Device Problem		
					CT3006	Implantable inorganic material	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem		
							F1901	Additional Surgery	A24	No Apparent Device Problem		
					CT314	Electrically powered	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem		
							F1901	Additional Surgery	A24	No Apparent Device Problem		
					CT321	Long-term surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem		
							F1901	Additional Surgery	A24	No Apparent Device Problem		
					CT449	Generators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem		
							F1901	Additional Surgery	A24	No Apparent Device Problem		
					CT981	Single use	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem		
							F1901	Additional Surgery	A24	No Apparent Device Problem		
					CT983	Surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem		
							F1901	Additional Surgery	A24	No Apparent Device Problem		
13048	2023	All other reportable incidents	35223	Endocardial/interventricular septal pacing lead	CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A0401	Break		
							F26	No Health Consequences or Impact	A0401	Break		
13049	2023	All other reportable incidents	35223	Endocardial/interventricular septal pacing lead	CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A05	Mechanical Problem		
									A0701	Capturing Problem		
									A072202	Low Impedance		
							F26	No Health Consequences or Impact	A05	Mechanical Problem		
									A0701	Capturing Problem		
									A072202	Low Impedance		
13050	2023	All other reportable incidents	35223	Endocardial/interventricular septal pacing lead	CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A072202	Low Impedance		
							F25	Unanticipated Adverse Device Effect	A072202	Low Impedance		
13051	2023	All other reportable incidents	17882	Physiologic-monitoring defibrillation system	CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A07	Electrical /Electronic Property Problem		
							F27	Problem Identified During Non-Clinical Procedure	A07	Electrical /Electronic Property Problem		
13052	2023	All other reportable incidents	17882	Physiologic-monitoring defibrillation system	CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070912	Incorrect Interpretation of Signal		
							F26	No Health Consequences or Impact	A070912	Incorrect Interpretation of Signal		
13053	2023	All other reportable incidents	35852	Single-chamber implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem	24/07/2023	24/07/2023
									A070504	Premature Discharge of Battery	24/07/2023	24/07/2023
							F08	Hospitalization or Prolonged Hospitalization	A0705	Battery Problem	24/07/2023	24/07/2023
									A070504	Premature Discharge of Battery	24/07/2023	24/07/2023
							F1905	Device Revision or Replacement	A0705	Battery Problem	24/07/2023	24/07/2023
									A070504	Premature Discharge of Battery	24/07/2023	24/07/2023
13054	2023	All other reportable incidents	37265	Dual-chamber implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A070912	Incorrect Interpretation of Signal		
									A071201	Failure to Convert Rhythm		
									A071302	Inappropriate/Inadequate Shock/Stimulation		
							F23	Unexpected Medical Intervention	A070912	Incorrect Interpretation of Signal		
									A071201	Failure to Convert Rhythm		
									A071302	Inappropriate/Inadequate Shock/Stimulation		
13055	2023	All other reportable incidents	36241	Pacemaker/defibrillator lead	CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070909	Oversensing		
									A090801	Signal Artifact/Noise		
							F26	No Health Consequences or Impact	A070909	Oversensing		
									A090801	Signal Artifact/Noise		
13056	2023		47265	Dual-chamber implantable pacemaker, rate-responsive	CT141	Electrodes/Leads and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A25	No Apparent Adverse Event		
							F26	No Health Consequences or Impact	A25	No Apparent Adverse Event		
					CT145	Cardiovascular devices	E2403	No Clinical Signs, Symptoms or Conditions	A25	No Apparent Adverse Event		
							F26	No Health Consequences or Impact	A25	No Apparent Adverse Event		
					CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A25	No Apparent Adverse Event		
							F26	No Health Consequences or Impact	A25	No Apparent Adverse Event		
					CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A25	No Apparent Adverse Event		
							F26	No Health Consequences or Impact	A25	No Apparent Adverse Event		
					CT233	Surgical	E2403	No Clinical Signs, Symptoms or Conditions	A25	No Apparent Adverse Event		
							F26	No Health Consequences or Impact	A25	No Apparent Adverse Event		
					CT2406	Implantable	E2403	No Clinical Signs, Symptoms or Conditions	A25	No Apparent Adverse Event		
							F26	No Health Consequences or Impact	A25	No Apparent Adverse Event		
					CT2407	Active implantable	E2403	No Clinical Signs, Symptoms or Conditions	A25	No Apparent Adverse Event		
							F26	No Health Consequences or Impact	A25	No Apparent Adverse Event		
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A25	No Apparent Adverse Event		
							F26	No Health Consequences or Impact	A25	No Apparent Adverse Event		
					CT2727	Active implant devices/ancillaries	E2403	No Clinical Signs, Symptoms or Conditions	A25	No Apparent Adverse Event		
							F26	No Health Consequences or Impact	A25	No Apparent Adverse Event		
					CT2750	Implantable cardiac electrical stimulators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A25	No Apparent Adverse Event		
							F26	No Health Consequences or Impact	A25	No Apparent Adverse Event		

					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A25	No Apparent Adverse Event		
							F26	No Health Consequences or Impact	A25	No Apparent Adverse Event		
					CT2985	Active	E2403	No Clinical Signs, Symptoms or Conditions	A25	No Apparent Adverse Event		
							F26	No Health Consequences or Impact	A25	No Apparent Adverse Event		
					CT3001	Implantable metal	E2403	No Clinical Signs, Symptoms or Conditions	A25	No Apparent Adverse Event		
							F26	No Health Consequences or Impact	A25	No Apparent Adverse Event		
					CT3002	Implantable synthetic polymer	E2403	No Clinical Signs, Symptoms or Conditions	A25	No Apparent Adverse Event		
							F26	No Health Consequences or Impact	A25	No Apparent Adverse Event		
					CT3006	Implantable inorganic material	E2403	No Clinical Signs, Symptoms or Conditions	A25	No Apparent Adverse Event		
							F26	No Health Consequences or Impact	A25	No Apparent Adverse Event		
					CT314	Electrically powered	E2403	No Clinical Signs, Symptoms or Conditions	A25	No Apparent Adverse Event		
							F26	No Health Consequences or Impact	A25	No Apparent Adverse Event		
					CT321	Long-term surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A25	No Apparent Adverse Event		
							F26	No Health Consequences or Impact	A25	No Apparent Adverse Event		
					CT449	Generators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A25	No Apparent Adverse Event		
							F26	No Health Consequences or Impact	A25	No Apparent Adverse Event		
					CT981	Single use	E2403	No Clinical Signs, Symptoms or Conditions	A25	No Apparent Adverse Event		
							F26	No Health Consequences or Impact	A25	No Apparent Adverse Event		
					CT983	Surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A25	No Apparent Adverse Event		
							F26	No Health Consequences or Impact	A25	No Apparent Adverse Event		
13057	2023	All other reportable incidents	47267	Single-chamber implantable pacemaker, rate-responsive	CT141	Electrodes/Leads and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	01/02/2023	01/02/2023
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	01/02/2023	01/02/2023
							F1903	Device Explantation	A070504	Premature Discharge of Battery	01/02/2023	01/02/2023
					CT145	Cardiovascular devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	01/02/2023	01/02/2023
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	01/02/2023	01/02/2023
							F1903	Device Explantation	A070504	Premature Discharge of Battery	01/02/2023	01/02/2023
					CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	01/02/2023	01/02/2023
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	01/02/2023	01/02/2023
							F1903	Device Explantation	A070504	Premature Discharge of Battery	01/02/2023	01/02/2023
					CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	01/02/2023	01/02/2023
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	01/02/2023	01/02/2023
							F1903	Device Explantation	A070504	Premature Discharge of Battery	01/02/2023	01/02/2023
					CT233	Surgical	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	01/02/2023	01/02/2023
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	01/02/2023	01/02/2023
							F1903	Device Explantation	A070504	Premature Discharge of Battery	01/02/2023	01/02/2023
					CT2406	Implantable	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	01/02/2023	01/02/2023
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	01/02/2023	01/02/2023
							F1903	Device Explantation	A070504	Premature Discharge of Battery	01/02/2023	01/02/2023
					CT2407	Active implantable	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	01/02/2023	01/02/2023
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	01/02/2023	01/02/2023
							F1903	Device Explantation	A070504	Premature Discharge of Battery	01/02/2023	01/02/2023
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	01/02/2023	01/02/2023
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	01/02/2023	01/02/2023
							F1903	Device Explantation	A070504	Premature Discharge of Battery	01/02/2023	01/02/2023
					CT2727	Active implant devices/ancillaries	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	01/02/2023	01/02/2023
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	01/02/2023	01/02/2023
							F1903	Device Explantation	A070504	Premature Discharge of Battery	01/02/2023	01/02/2023
					CT2750	Implantable cardiac electrical stimulators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	01/02/2023	01/02/2023
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	01/02/2023	01/02/2023
							F1903	Device Explantation	A070504	Premature Discharge of Battery	01/02/2023	01/02/2023
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	01/02/2023	01/02/2023
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	01/02/2023	01/02/2023
							F1903	Device Explantation	A070504	Premature Discharge of Battery	01/02/2023	01/02/2023
					CT2985	Active	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	01/02/2023	01/02/2023
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	01/02/2023	01/02/2023
							F1903	Device Explantation	A070504	Premature Discharge of Battery	01/02/2023	01/02/2023
					CT3001	Implantable metal	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	01/02/2023	01/02/2023
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	01/02/2023	01/02/2023
							F1903	Device Explantation	A070504	Premature Discharge of Battery	01/02/2023	01/02/2023
					CT3002	Implantable synthetic polymer	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	01/02/2023	01/02/2023
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	01/02/2023	01/02/2023
							F1903	Device Explantation	A070504	Premature Discharge of Battery	01/02/2023	01/02/2023
					CT3006	Implantable inorganic material	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	01/02/2023	01/02/2023
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	01/02/2023	01/02/2023
							F1903	Device Explantation	A070504	Premature Discharge of Battery	01/02/2023	01/02/2023
					CT314	Electrically powered	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	01/02/2023	01/02/2023
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	01/02/2023	01/02/2023
							F1903	Device Explantation	A070504	Premature Discharge of Battery	01/02/2023	01/02/2023
					CT321	Long-term surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	01/02/2023	01/02/2023
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	01/02/2023	01/02/2023
							F1903	Device Explantation	A070504	Premature Discharge of Battery	01/02/2023	01/02/2023
					CT449	Generators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	01/02/2023	01/02/2023
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	01/02/2023	01/02/2023
							F1903	Device Explantation	A070504	Premature Discharge of Battery	01/02/2023	01/02/2023
					CT981	Single use	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	01/02/2023	01/02/2023
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	01/02/2023	01/02/2023
							F1903	Device Explantation	A070504	Premature Discharge of Battery	01/02/2023	01/02/2023
					CT983	Surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	01/02/2023	01/02/2023
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	01/02/2023	01/02/2023
							F1903	Device Explantation	A070504	Premature Discharge of Battery	01/02/2023	01/02/2023
13058	2023	All other reportable incidents	47205	Cardiac pulse generator programmer	CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A0203	Defective Device		
									A150105	Key or Button Unresponsive/not Working		
							F26	No Health Consequences or Impact	A0203	Defective Device		
									A150105	Key or Button Unresponsive/not Working		
13059	2023	All other reportable incidents	47265	Dual-chamber implantable pacemaker, rate-responsive	CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem	23/06/2023	23/06/2023
									A070504	Premature Discharge of Battery	23/06/2023	23/06/2023
							F08	Hospitalization or Prolonged Hospitalization	A0705	Battery Problem	23/06/2023	23/06/2023
									A070504	Premature Discharge of Battery	23/06/2023	23/06/2023

							F1905	Device Revision or Replacement	A0705	Battery Problem	23/06/2023	23/06/2023
									A070504	Premature Discharge of Battery	23/06/2023	23/06/2023
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem	23/06/2023	23/06/2023
									A070504	Premature Discharge of Battery	23/06/2023	23/06/2023
							F08	Hospitalization or Prolonged Hospitalization	A0705	Battery Problem	23/06/2023	23/06/2023
									A070504	Premature Discharge of Battery	23/06/2023	23/06/2023
							F1905	Device Revision or Replacement	A0705	Battery Problem	23/06/2023	23/06/2023
									A070504	Premature Discharge of Battery	23/06/2023	23/06/2023
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem	23/06/2023	23/06/2023
									A070504	Premature Discharge of Battery	23/06/2023	23/06/2023
							F08	Hospitalization or Prolonged Hospitalization	A0705	Battery Problem	23/06/2023	23/06/2023
									A070504	Premature Discharge of Battery	23/06/2023	23/06/2023
							F1905	Device Revision or Replacement	A0705	Battery Problem	23/06/2023	23/06/2023
									A070504	Premature Discharge of Battery	23/06/2023	23/06/2023
13060	2023	All other reportable incidents	47263	Cardiac resynchronization therapy implantable pacemaker	CT1685	Cardiac pacemakers	E2103	Device Overstimulation of Tissue	A0908	Incorrect, Inadequate or Imprecise Result or Readings	16/08/2022	16/08/2022
									A0909	Unexpected Therapeutic Results	16/08/2022	16/08/2022
							F08	Hospitalization or Prolonged Hospitalization	A0908	Incorrect, Inadequate or Imprecise Result or Readings	16/08/2022	16/08/2022
									A0909	Unexpected Therapeutic Results	16/08/2022	16/08/2022
							F1905	Device Revision or Replacement	A0908	Incorrect, Inadequate or Imprecise Result or Readings	16/08/2022	16/08/2022
									A0909	Unexpected Therapeutic Results	16/08/2022	16/08/2022
					CT257	Implantable cardiac pacemakers	E2103	Device Overstimulation of Tissue	A0908	Incorrect, Inadequate or Imprecise Result or Readings	16/08/2022	16/08/2022
									A0909	Unexpected Therapeutic Results	16/08/2022	16/08/2022
							F08	Hospitalization or Prolonged Hospitalization	A0908	Incorrect, Inadequate or Imprecise Result or Readings	16/08/2022	16/08/2022
									A0909	Unexpected Therapeutic Results	16/08/2022	16/08/2022
							F1905	Device Revision or Replacement	A0908	Incorrect, Inadequate or Imprecise Result or Readings	16/08/2022	16/08/2022
									A0909	Unexpected Therapeutic Results	16/08/2022	16/08/2022
					CT2866	Implantable cardiac electrical stimulators	E2103	Device Overstimulation of Tissue	A0908	Incorrect, Inadequate or Imprecise Result or Readings	16/08/2022	16/08/2022
									A0909	Unexpected Therapeutic Results	16/08/2022	16/08/2022
							F08	Hospitalization or Prolonged Hospitalization	A0908	Incorrect, Inadequate or Imprecise Result or Readings	16/08/2022	16/08/2022
									A0909	Unexpected Therapeutic Results	16/08/2022	16/08/2022
							F1905	Device Revision or Replacement	A0908	Incorrect, Inadequate or Imprecise Result or Readings	16/08/2022	16/08/2022
									A0909	Unexpected Therapeutic Results	16/08/2022	16/08/2022
13061	2023	All other reportable incidents	17882	Physiologic-monitoring defibrillation system	CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A0706	Charging Problem		
							F27	Problem Identified During Non-Clinical Procedure	A0706	Charging Problem		
13062	2023	All other reportable incidents	47270	Cardiac resynchronization therapy implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A0713	Defibrillation/Stimulation Problem		
							F26	No Health Consequences or Impact	A0713	Defibrillation/Stimulation Problem		
13063	2023	All other reportable incidents	35852	Single-chamber implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A070909	Oversensing		
									A0712	Pacing Problem		
							F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A070909	Oversensing		
									A0712	Pacing Problem		
13064	2023	All other reportable incidents	35852	Single-chamber implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E2104	Electric Shock	A0709	Device Sensing Problem		
									A070909	Oversensing		
									A071302	Inappropriate/Inadequate Shock/Stimulation		
							E2330	Pain	A0709	Device Sensing Problem		
									A070909	Oversensing		
									A071302	Inappropriate/Inadequate Shock/Stimulation		
							F08	Hospitalization or Prolonged Hospitalization	A0709	Device Sensing Problem		
									A070909	Oversensing		
									A071302	Inappropriate/Inadequate Shock/Stimulation		
							F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A0709	Device Sensing Problem		
									A070909	Oversensing		
									A071302	Inappropriate/Inadequate Shock/Stimulation		
							F2203	Imaging Required	A0709	Device Sensing Problem		
									A070909	Oversensing		
									A071302	Inappropriate/Inadequate Shock/Stimulation		
13065	2023		47265	Dual-chamber implantable pacemaker, rate-responsive	CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	28/04/2023	28/04/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	28/04/2023	28/04/2023
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	28/04/2023	28/04/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	28/04/2023	28/04/2023
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	28/04/2023	28/04/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	28/04/2023	28/04/2023
13066	2023		47265	Dual-chamber implantable pacemaker, rate-responsive	CT141	Electrodes/Leads and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	12/02/2023	12/02/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	12/02/2023	12/02/2023
					CT145	Cardiovascular devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	12/02/2023	12/02/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	12/02/2023	12/02/2023
					CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	12/02/2023	12/02/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	12/02/2023	12/02/2023
					CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	12/02/2023	12/02/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	12/02/2023	12/02/2023
					CT233	Surgical	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	12/02/2023	12/02/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	12/02/2023	12/02/2023
					CT2406	Implantable	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	12/02/2023	12/02/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	12/02/2023	12/02/2023
					CT2407	Active implantable	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	12/02/2023	12/02/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	12/02/2023	12/02/2023
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	12/02/2023	12/02/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	12/02/2023	12/02/2023
					CT2727	Active implant devices/ancillaries	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	12/02/2023	12/02/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	12/02/2023	12/02/2023
					CT2750	Implantable cardiac electrical stimulators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	12/02/2023	12/02/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	12/02/2023	12/02/2023

					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	12/02/2023	12/02/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	12/02/2023	12/02/2023
					CT2985	Active	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	12/02/2023	12/02/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	12/02/2023	12/02/2023
					CT3001	Implantable metal	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	12/02/2023	12/02/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	12/02/2023	12/02/2023
					CT3002	Implantable synthetic polymer	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	12/02/2023	12/02/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	12/02/2023	12/02/2023
					CT3006	Implantable inorganic material	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	12/02/2023	12/02/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	12/02/2023	12/02/2023
					CT314	Electrically powered	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	12/02/2023	12/02/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	12/02/2023	12/02/2023
					CT321	Long-term surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	12/02/2023	12/02/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	12/02/2023	12/02/2023
					CT449	Generators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	12/02/2023	12/02/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	12/02/2023	12/02/2023
					CT981	Single use	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	12/02/2023	12/02/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	12/02/2023	12/02/2023
					CT983	Surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	12/02/2023	12/02/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	12/02/2023	12/02/2023
13067	2023		47265	Dual-chamber implantable pacemaker, rate-responsive	CT141	Electrodes/Leads and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	19/02/2023	19/02/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	19/02/2023	19/02/2023
					CT145	Cardiovascular devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	19/02/2023	19/02/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	19/02/2023	19/02/2023
					CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	19/02/2023	19/02/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	19/02/2023	19/02/2023
					CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	19/02/2023	19/02/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	19/02/2023	19/02/2023
					CT233	Surgical	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	19/02/2023	19/02/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	19/02/2023	19/02/2023
					CT2406	Implantable	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	19/02/2023	19/02/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	19/02/2023	19/02/2023
					CT2407	Active implantable	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	19/02/2023	19/02/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	19/02/2023	19/02/2023
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	19/02/2023	19/02/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	19/02/2023	19/02/2023
					CT2727	Active implant devices/ancillaries	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	19/02/2023	19/02/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	19/02/2023	19/02/2023
					CT2750	Implantable cardiac electrical stimulators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	19/02/2023	19/02/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	19/02/2023	19/02/2023
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	19/02/2023	19/02/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	19/02/2023	19/02/2023
					CT2985	Active	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	19/02/2023	19/02/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	19/02/2023	19/02/2023
					CT3001	Implantable metal	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	19/02/2023	19/02/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	19/02/2023	19/02/2023
					CT3002	Implantable synthetic polymer	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	19/02/2023	19/02/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	19/02/2023	19/02/2023
					CT3006	Implantable inorganic material	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	19/02/2023	19/02/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	19/02/2023	19/02/2023
					CT314	Electrically powered	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	19/02/2023	19/02/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	19/02/2023	19/02/2023
					CT321	Long-term surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	19/02/2023	19/02/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	19/02/2023	19/02/2023
					CT449	Generators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	19/02/2023	19/02/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	19/02/2023	19/02/2023
					CT981	Single use	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	19/02/2023	19/02/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	19/02/2023	19/02/2023
					CT983	Surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	19/02/2023	19/02/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	19/02/2023	19/02/2023
13068	2023	All other reportable incidents	37265	Dual-chamber implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	22/11/2022	22/11/2022
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	22/11/2022	22/11/2022
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	22/11/2022	22/11/2022
13069	2023		47265	Dual-chamber implantable pacemaker, rate-responsive	CT141	Electrodes/Leads and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	12/02/2023	12/02/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	12/02/2023	12/02/2023
					CT145	Cardiovascular devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	12/02/2023	12/02/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	12/02/2023	12/02/2023
					CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	12/02/2023	12/02/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	12/02/2023	12/02/2023
					CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	12/02/2023	12/02/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	12/02/2023	12/02/2023
					CT233	Surgical	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	12/02/2023	12/02/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	12/02/2023	12/02/2023
					CT2406	Implantable	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	12/02/2023	12/02/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	12/02/2023	12/02/2023
					CT2407	Active implantable	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	12/02/2023	12/02/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	12/02/2023	12/02/2023
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	12/02/2023	12/02/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	12/02/2023	12/02/2023
					CT2727	Active implant devices/ancillaries	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	12/02/2023	12/02/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	12/02/2023	12/02/2023
					CT2750	Implantable cardiac electrical stimulators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	12/02/2023	12/02/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	12/02/2023	12/02/2023
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	12/02/2023	12/02/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	12/02/2023	12/02/2023
					CT2985	Active	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	12/02/2023	12/02/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	12/02/2023	12/02/2023
					CT3001	Implantable metal	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	12/02/2023	12/02/2023

							F1901	Additional Surgery	A24	No Apparent Device Problem	12/02/2023	12/02/2023
					CT3002	Implantable synthetic polymer	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	12/02/2023	12/02/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	12/02/2023	12/02/2023
					CT3006	Implantable inorganic material	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	12/02/2023	12/02/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	12/02/2023	12/02/2023
					CT314	Electrically powered	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	12/02/2023	12/02/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	12/02/2023	12/02/2023
					CT321	Long-term surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	12/02/2023	12/02/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	12/02/2023	12/02/2023
					CT449	Generators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	12/02/2023	12/02/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	12/02/2023	12/02/2023
					CT981	Single use	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	12/02/2023	12/02/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	12/02/2023	12/02/2023
					CT983	Surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	12/02/2023	12/02/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	12/02/2023	12/02/2023
13070	2023	All other reportable incidents	35223	Endocardial/interventricular septal pacing lead	CT1684	Cardiac pacemakers and associated devices	E060109	Tachycardia	A070909	Oversensing		
							F1001	Absence of Treatment	A070909	Oversensing		
							F1203	Life Threatening Illness or Injury	A070909	Oversensing		
							F23	Unexpected Medical Intervention	A070909	Oversensing		
13071	2023	All other reportable incidents	47265	Dual-chamber implantable pacemaker, rate-responsive	CT141	Electrodes/Leads and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	23/03/2023	23/03/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	23/03/2023	23/03/2023
					CT145	Cardiovascular devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	23/03/2023	23/03/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	23/03/2023	23/03/2023
					CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	23/03/2023	23/03/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	23/03/2023	23/03/2023
					CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	23/03/2023	23/03/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	23/03/2023	23/03/2023
					CT233	Surgical	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	23/03/2023	23/03/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	23/03/2023	23/03/2023
					CT2406	Implantable	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	23/03/2023	23/03/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	23/03/2023	23/03/2023
					CT2407	Active implantable	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	23/03/2023	23/03/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	23/03/2023	23/03/2023
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	23/03/2023	23/03/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	23/03/2023	23/03/2023
					CT2727	Active implant devices/ancillaries	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	23/03/2023	23/03/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	23/03/2023	23/03/2023
					CT2750	Implantable cardiac electrical stimulators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	23/03/2023	23/03/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	23/03/2023	23/03/2023
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	23/03/2023	23/03/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	23/03/2023	23/03/2023
					CT2985	Active	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	23/03/2023	23/03/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	23/03/2023	23/03/2023
					CT3001	Implantable metal	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	23/03/2023	23/03/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	23/03/2023	23/03/2023
					CT3002	Implantable synthetic polymer	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	23/03/2023	23/03/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	23/03/2023	23/03/2023
					CT3006	Implantable inorganic material	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	23/03/2023	23/03/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	23/03/2023	23/03/2023
					CT314	Electrically powered	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	23/03/2023	23/03/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	23/03/2023	23/03/2023
					CT321	Long-term surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	23/03/2023	23/03/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	23/03/2023	23/03/2023
					CT449	Generators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	23/03/2023	23/03/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	23/03/2023	23/03/2023
					CT981	Single use	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	23/03/2023	23/03/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	23/03/2023	23/03/2023
					CT983	Surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	23/03/2023	23/03/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	23/03/2023	23/03/2023
13072	2023	All other reportable incidents	47265	Dual-chamber implantable pacemaker, rate-responsive	CT141	Electrodes/Leads and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	05/02/2023	05/02/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	05/02/2023	05/02/2023
					CT145	Cardiovascular devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	05/02/2023	05/02/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	05/02/2023	05/02/2023
					CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	05/02/2023	05/02/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	05/02/2023	05/02/2023
					CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	05/02/2023	05/02/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	05/02/2023	05/02/2023
					CT233	Surgical	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	05/02/2023	05/02/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	05/02/2023	05/02/2023
					CT2406	Implantable	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	05/02/2023	05/02/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	05/02/2023	05/02/2023
					CT2407	Active implantable	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	05/02/2023	05/02/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	05/02/2023	05/02/2023
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	05/02/2023	05/02/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	05/02/2023	05/02/2023
					CT2727	Active implant devices/ancillaries	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	05/02/2023	05/02/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	05/02/2023	05/02/2023
					CT2750	Implantable cardiac electrical stimulators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	05/02/2023	05/02/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	05/02/2023	05/02/2023
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	05/02/2023	05/02/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	05/02/2023	05/02/2023
					CT2985	Active	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	05/02/2023	05/02/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	05/02/2023	05/02/2023
					CT3001	Implantable metal	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	05/02/2023	05/02/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	05/02/2023	05/02/2023
					CT3002	Implantable synthetic polymer	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	05/02/2023	05/02/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	05/02/2023	05/02/2023
					CT3006	Implantable inorganic material	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	05/02/2023	05/02/2023

						F1901	Additional Surgery	A24	No Apparent Device Problem	05/02/2023	05/02/2023
				CT314	Electrically powered	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	05/02/2023	05/02/2023
						F1901	Additional Surgery	A24	No Apparent Device Problem	05/02/2023	05/02/2023
				CT321	Long-term surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	05/02/2023	05/02/2023
						F1901	Additional Surgery	A24	No Apparent Device Problem	05/02/2023	05/02/2023
				CT449	Generators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	05/02/2023	05/02/2023
						F1901	Additional Surgery	A24	No Apparent Device Problem	05/02/2023	05/02/2023
				CT981	Single use	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	05/02/2023	05/02/2023
						F1901	Additional Surgery	A24	No Apparent Device Problem	05/02/2023	05/02/2023
				CT983	Surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	05/02/2023	05/02/2023
						F1901	Additional Surgery	A24	No Apparent Device Problem	05/02/2023	05/02/2023
13073	2023		47265	Dual-chamber implantable pacemaker, rate-responsive	CT141	Electrodes/Leads and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	11/04/2023	11/04/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	11/04/2023	11/04/2023
				CT145	Cardiovascular devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	11/04/2023	11/04/2023
						F1901	Additional Surgery	A24	No Apparent Device Problem	11/04/2023	11/04/2023
				CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	11/04/2023	11/04/2023
						F1901	Additional Surgery	A24	No Apparent Device Problem	11/04/2023	11/04/2023
				CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	11/04/2023	11/04/2023
						F1901	Additional Surgery	A24	No Apparent Device Problem	11/04/2023	11/04/2023
				CT233	Surgical	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	11/04/2023	11/04/2023
						F1901	Additional Surgery	A24	No Apparent Device Problem	11/04/2023	11/04/2023
				CT2406	Implantable	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	11/04/2023	11/04/2023
						F1901	Additional Surgery	A24	No Apparent Device Problem	11/04/2023	11/04/2023
				CT2407	Active implantable	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	11/04/2023	11/04/2023
						F1901	Additional Surgery	A24	No Apparent Device Problem	11/04/2023	11/04/2023
				CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	11/04/2023	11/04/2023
						F1901	Additional Surgery	A24	No Apparent Device Problem	11/04/2023	11/04/2023
				CT2727	Active implant devices/ancillaries	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	11/04/2023	11/04/2023
						F1901	Additional Surgery	A24	No Apparent Device Problem	11/04/2023	11/04/2023
				CT2750	Implantable cardiac electrical stimulators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	11/04/2023	11/04/2023
						F1901	Additional Surgery	A24	No Apparent Device Problem	11/04/2023	11/04/2023
				CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	11/04/2023	11/04/2023
						F1901	Additional Surgery	A24	No Apparent Device Problem	11/04/2023	11/04/2023
				CT2985	Active	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	11/04/2023	11/04/2023
						F1901	Additional Surgery	A24	No Apparent Device Problem	11/04/2023	11/04/2023
				CT3001	Implantable metal	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	11/04/2023	11/04/2023
						F1901	Additional Surgery	A24	No Apparent Device Problem	11/04/2023	11/04/2023
				CT3002	Implantable synthetic polymer	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	11/04/2023	11/04/2023
						F1901	Additional Surgery	A24	No Apparent Device Problem	11/04/2023	11/04/2023
				CT3006	Implantable inorganic material	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	11/04/2023	11/04/2023
						F1901	Additional Surgery	A24	No Apparent Device Problem	11/04/2023	11/04/2023
				CT314	Electrically powered	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	11/04/2023	11/04/2023
						F1901	Additional Surgery	A24	No Apparent Device Problem	11/04/2023	11/04/2023
				CT321	Long-term surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	11/04/2023	11/04/2023
						F1901	Additional Surgery	A24	No Apparent Device Problem	11/04/2023	11/04/2023
				CT449	Generators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	11/04/2023	11/04/2023
						F1901	Additional Surgery	A24	No Apparent Device Problem	11/04/2023	11/04/2023
				CT981	Single use	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	11/04/2023	11/04/2023
						F1901	Additional Surgery	A24	No Apparent Device Problem	11/04/2023	11/04/2023
				CT983	Surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	11/04/2023	11/04/2023
						F1901	Additional Surgery	A24	No Apparent Device Problem	11/04/2023	11/04/2023
13074	2023	All other reportable incidents	35852	Single-chamber implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	03/08/2023	03/08/2023
									A0908	Incorrect, Inadequate or Imprecise Result or Readings	03/08/2023	03/08/2023
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	03/08/2023	03/08/2023
									A0908	Incorrect, Inadequate or Imprecise Result or Readings	03/08/2023	03/08/2023
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	03/08/2023	03/08/2023
									A0908	Incorrect, Inadequate or Imprecise Result or Readings	03/08/2023	03/08/2023
13075	2023	All other reportable incidents	35852	Single-chamber implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A160602	Premature End-of-Life Indicator		
							F26	No Health Consequences or Impact	A160602	Premature End-of-Life Indicator		
13076	2023	All other reportable incidents	17882	Physiologic-monitoring defibrillation system	CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A1603	Failure of Device to Self-Test		
							F26	No Health Consequences or Impact	A1603	Failure of Device to Self-Test		
13077	2023	All other reportable incidents	37265	Dual-chamber implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A07	Electrical /Electronic Property Problem	05/05/2023	05/05/2023
									A070504	Premature Discharge of Battery	05/05/2023	05/05/2023
							F08	Hospitalization or Prolonged Hospitalization	A07	Electrical /Electronic Property Problem	05/05/2023	05/05/2023
									A070504	Premature Discharge of Battery	05/05/2023	05/05/2023
							F12	Serious Injury/ Illness/ Impairment	A07	Electrical /Electronic Property Problem	05/05/2023	05/05/2023
									A070504	Premature Discharge of Battery	05/05/2023	05/05/2023
							F1905	Device Revision or Replacement	A07	Electrical /Electronic Property Problem	05/05/2023	05/05/2023
									A070504	Premature Discharge of Battery	05/05/2023	05/05/2023
13078	2023		47265	Dual-chamber implantable pacemaker, rate-responsive	CT141	Electrodes/Leads and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	14/03/2023	14/03/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	14/03/2023	14/03/2023
				CT145	Cardiovascular devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	14/03/2023	14/03/2023
						F1901	Additional Surgery	A24	No Apparent Device Problem	14/03/2023	14/03/2023
				CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	14/03/2023	14/03/2023
						F1901	Additional Surgery	A24	No Apparent Device Problem	14/03/2023	14/03/2023
				CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	14/03/2023	14/03/2023
						F1901	Additional Surgery	A24	No Apparent Device Problem	14/03/2023	14/03/2023
				CT233	Surgical	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	14/03/2023	14/03/2023
						F1901	Additional Surgery	A24	No Apparent Device Problem	14/03/2023	14/03/2023
				CT2406	Implantable	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	14/03/2023	14/03/2023
						F1901	Additional Surgery	A24	No Apparent Device Problem	14/03/2023	14/03/2023
				CT2407	Active implantable	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	14/03/2023	14/03/2023
						F1901	Additional Surgery	A24	No Apparent Device Problem	14/03/2023	14/03/2023
				CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	14/03/2023	14/03/2023
						F1901	Additional Surgery	A24	No Apparent Device Problem	14/03/2023	14/03/2023

					CT2727	Active implant devices/ancillaries	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	14/03/2023	14/03/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	14/03/2023	14/03/2023
					CT2750	Implantable cardiac electrical stimulators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	14/03/2023	14/03/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	14/03/2023	14/03/2023
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	14/03/2023	14/03/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	14/03/2023	14/03/2023
					CT2985	Active	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	14/03/2023	14/03/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	14/03/2023	14/03/2023
					CT3001	Implantable metal	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	14/03/2023	14/03/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	14/03/2023	14/03/2023
					CT3002	Implantable synthetic polymer	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	14/03/2023	14/03/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	14/03/2023	14/03/2023
					CT3006	Implantable inorganic material	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	14/03/2023	14/03/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	14/03/2023	14/03/2023
					CT314	Electrically powered	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	14/03/2023	14/03/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	14/03/2023	14/03/2023
					CT321	Long-term surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	14/03/2023	14/03/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	14/03/2023	14/03/2023
					CT449	Generators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	14/03/2023	14/03/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	14/03/2023	14/03/2023
					CT981	Single use	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	14/03/2023	14/03/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	14/03/2023	14/03/2023
					CT983	Surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	14/03/2023	14/03/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	14/03/2023	14/03/2023
13079	2023	All other reportable incidents	47265	Dual-chamber implantable pacemaker, rate-responsive	CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	21/03/2023	21/03/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	21/03/2023	21/03/2023
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	21/03/2023	21/03/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	21/03/2023	21/03/2023
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	21/03/2023	21/03/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	21/03/2023	21/03/2023
13080	2023	Death	47270	Cardiac resynchronization therapy implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E0602	Cardiac Arrest	A24	No Apparent Device Problem		
							F02	Death	A24	No Apparent Device Problem		
13081	2023	All other reportable incidents	47265	Dual-chamber implantable pacemaker, rate-responsive	CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A071102	Failure to Interrogate	09/05/2023	09/05/2023
							F1901	Additional Surgery	A071102	Failure to Interrogate	09/05/2023	09/05/2023
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A071102	Failure to Interrogate	09/05/2023	09/05/2023
							F1901	Additional Surgery	A071102	Failure to Interrogate	09/05/2023	09/05/2023
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A071102	Failure to Interrogate	09/05/2023	09/05/2023
							F1901	Additional Surgery	A071102	Failure to Interrogate	09/05/2023	09/05/2023
13082	2023	All other reportable incidents	47205	Cardiac pulse generator programmer	CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A150105	Key or Button Unresponsive/not Working		
							F26	No Health Consequences or Impact	A150105	Key or Button Unresponsive/not Working		
13083	2023		47265	Dual-chamber implantable pacemaker, rate-responsive	CT141	Electrodes/Leads and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	07/03/2023	07/03/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	07/03/2023	07/03/2023
					CT145	Cardiovascular devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	07/03/2023	07/03/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	07/03/2023	07/03/2023
					CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	07/03/2023	07/03/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	07/03/2023	07/03/2023
					CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	07/03/2023	07/03/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	07/03/2023	07/03/2023
					CT233	Surgical	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	07/03/2023	07/03/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	07/03/2023	07/03/2023
					CT2406	Implantable	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	07/03/2023	07/03/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	07/03/2023	07/03/2023
					CT2407	Active implantable	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	07/03/2023	07/03/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	07/03/2023	07/03/2023
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	07/03/2023	07/03/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	07/03/2023	07/03/2023
					CT2727	Active implant devices/ancillaries	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	07/03/2023	07/03/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	07/03/2023	07/03/2023
					CT2750	Implantable cardiac electrical stimulators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	07/03/2023	07/03/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	07/03/2023	07/03/2023
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	07/03/2023	07/03/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	07/03/2023	07/03/2023
					CT2985	Active	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	07/03/2023	07/03/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	07/03/2023	07/03/2023
					CT3001	Implantable metal	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	07/03/2023	07/03/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	07/03/2023	07/03/2023
					CT3002	Implantable synthetic polymer	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	07/03/2023	07/03/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	07/03/2023	07/03/2023
					CT3006	Implantable inorganic material	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	07/03/2023	07/03/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	07/03/2023	07/03/2023
					CT314	Electrically powered	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	07/03/2023	07/03/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	07/03/2023	07/03/2023
					CT321	Long-term surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	07/03/2023	07/03/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	07/03/2023	07/03/2023
					CT449	Generators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	07/03/2023	07/03/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	07/03/2023	07/03/2023
					CT981	Single use	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	07/03/2023	07/03/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	07/03/2023	07/03/2023
					CT983	Surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	07/03/2023	07/03/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	07/03/2023	07/03/2023
13084	2023	All other reportable incidents	35223	Endocardial/interventricular septal pacing lead	CT1684	Cardiac pacemakers and associated devices	E0402	Hypersensitivity/Allergic Reaction	A24	No Apparent Device Problem	01/04/2023	01/04/2023
							F1903	Device Explantation	A24	No Apparent Device Problem	01/04/2023	01/04/2023
13085	2023	All other reportable incidents	37265	Dual-chamber implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E0402	Hypersensitivity/Allergic Reaction	A24	No Apparent Device Problem	01/04/2023	01/04/2023
							F1903	Device Explantation	A24	No Apparent Device Problem	01/04/2023	01/04/2023
13086	2023	All other reportable incidents	47205	Cardiac pulse generator programmer	CT1684	Cardiac pacemakers and associated devices	E060101	Asystole	A071102	Failure to Interrogate		
									A0902	Display or Visual Feedback Problem		
									A1104	Computer Operating System Problem		
									A13	Communication or Transmission Problem		

							F12	Serious Injury/ Illness/ Impairment	A071102	Failure to Interrogate		
									A0902	Display or Visual Feedback Problem		
									A1104	Computer Operating System Problem		
									A13	Communication or Transmission Problem		
							F23	Unexpected Medical Intervention	A071102	Failure to Interrogate		
									A0902	Display or Visual Feedback Problem		
									A1104	Computer Operating System Problem		
									A13	Communication or Transmission Problem		
13087	2023	All other reportable incidents	47267	Single-chamber implantable pacemaker, rate-responsive	CT141	Electrodes/Leads and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A071102	Failure to Interrogate		
							F26	No Health Consequences or Impact	A071102	Failure to Interrogate		
					CT145	Cardiovascular devices	E2403	No Clinical Signs, Symptoms or Conditions	A071102	Failure to Interrogate		
							F26	No Health Consequences or Impact	A071102	Failure to Interrogate		
					CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A071102	Failure to Interrogate		
							F26	No Health Consequences or Impact	A071102	Failure to Interrogate		
					CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A071102	Failure to Interrogate		
							F26	No Health Consequences or Impact	A071102	Failure to Interrogate		
					CT233	Surgical	E2403	No Clinical Signs, Symptoms or Conditions	A071102	Failure to Interrogate		
							F26	No Health Consequences or Impact	A071102	Failure to Interrogate		
					CT2406	Implantable	E2403	No Clinical Signs, Symptoms or Conditions	A071102	Failure to Interrogate		
							F26	No Health Consequences or Impact	A071102	Failure to Interrogate		
					CT2407	Active implantable	E2403	No Clinical Signs, Symptoms or Conditions	A071102	Failure to Interrogate		
							F26	No Health Consequences or Impact	A071102	Failure to Interrogate		
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A071102	Failure to Interrogate		
							F26	No Health Consequences or Impact	A071102	Failure to Interrogate		
					CT2727	Active implant devices/ancillaries	E2403	No Clinical Signs, Symptoms or Conditions	A071102	Failure to Interrogate		
							F26	No Health Consequences or Impact	A071102	Failure to Interrogate		
					CT2750	Implantable cardiac electrical stimulators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A071102	Failure to Interrogate		
							F26	No Health Consequences or Impact	A071102	Failure to Interrogate		
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A071102	Failure to Interrogate		
							F26	No Health Consequences or Impact	A071102	Failure to Interrogate		
					CT2985	Active	E2403	No Clinical Signs, Symptoms or Conditions	A071102	Failure to Interrogate		
							F26	No Health Consequences or Impact	A071102	Failure to Interrogate		
					CT3001	Implantable metal	E2403	No Clinical Signs, Symptoms or Conditions	A071102	Failure to Interrogate		
							F26	No Health Consequences or Impact	A071102	Failure to Interrogate		
					CT3002	Implantable synthetic polymer	E2403	No Clinical Signs, Symptoms or Conditions	A071102	Failure to Interrogate		
							F26	No Health Consequences or Impact	A071102	Failure to Interrogate		
					CT3006	Implantable inorganic material	E2403	No Clinical Signs, Symptoms or Conditions	A071102	Failure to Interrogate		
							F26	No Health Consequences or Impact	A071102	Failure to Interrogate		
					CT314	Electrically powered	E2403	No Clinical Signs, Symptoms or Conditions	A071102	Failure to Interrogate		
							F26	No Health Consequences or Impact	A071102	Failure to Interrogate		
					CT321	Long-term surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A071102	Failure to Interrogate		
							F26	No Health Consequences or Impact	A071102	Failure to Interrogate		
					CT449	Generators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A071102	Failure to Interrogate		
							F26	No Health Consequences or Impact	A071102	Failure to Interrogate		
					CT981	Single use	E2403	No Clinical Signs, Symptoms or Conditions	A071102	Failure to Interrogate		
							F26	No Health Consequences or Impact	A071102	Failure to Interrogate		
					CT983	Surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A071102	Failure to Interrogate		
							F26	No Health Consequences or Impact	A071102	Failure to Interrogate		
13088	2023	All other reportable incidents	35852	Single-chamber implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	13/01/2022	13/01/2022
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	13/01/2022	13/01/2022
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	13/01/2022	13/01/2022
13089	2023	All other reportable incidents	35852	Single-chamber implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A070909	Oversensing		
							F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A070909	Oversensing		
13090	2023	All other reportable incidents	47205	Cardiac pulse generator programmer	CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A150105	Key or Button Unresponsive/not Working		
							F27	Problem Identified During Non-Clinical Procedure	A150105	Key or Button Unresponsive/not Working		
13091	2023	All other reportable incidents	36241	Pacemaker/defibrillator lead	CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A072201	High Impedance		
							F26	No Health Consequences or Impact	A072201	High Impedance		
13092	2023	All other reportable incidents	47205	Cardiac pulse generator programmer	CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A150105	Key or Button Unresponsive/not Working		
							F1908	Prolonged Surgery	A150105	Key or Button Unresponsive/not Working		
13093	2023	All other reportable incidents	47263	Cardiac resynchronization therapy implantable pacemaker	CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	13/07/2022	13/07/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	13/07/2022	13/07/2022
					CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	13/07/2022	13/07/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	13/07/2022	13/07/2022
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	13/07/2022	13/07/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	13/07/2022	13/07/2022
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	13/07/2022	13/07/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	13/07/2022	13/07/2022
13094	2023	All other reportable incidents	47205	Cardiac pulse generator programmer	CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A090805	Incorrect Measurement		
							F23	Unexpected Medical Intervention	A090805	Incorrect Measurement		
13095	2023	All other reportable incidents	47205	Cardiac pulse generator programmer	CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A090805	Incorrect Measurement		
							F23	Unexpected Medical Intervention	A090805	Incorrect Measurement		
13096	2023		47263	Cardiac resynchronization therapy implantable pacemaker	CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A25	No Apparent Adverse Event	13/07/2022	13/07/2022
							F26	No Health Consequences or Impact	A25	No Apparent Adverse Event	13/07/2022	13/07/2022
					CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A25	No Apparent Adverse Event	13/07/2022	13/07/2022
							F26	No Health Consequences or Impact	A25	No Apparent Adverse Event	13/07/2022	13/07/2022
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A25	No Apparent Adverse Event	13/07/2022	13/07/2022
							F26	No Health Consequences or Impact	A25	No Apparent Adverse Event	13/07/2022	13/07/2022
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A25	No Apparent Adverse Event	13/07/2022	13/07/2022
							F26	No Health Consequences or Impact	A25	No Apparent Adverse Event	13/07/2022	13/07/2022
13097	2023		47205	Cardiac pulse generator programmer	CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A25	No Apparent Adverse Event		
							F26	No Health Consequences or Impact	A25	No Apparent Adverse Event		
13098	2023		47263	Cardiac resynchronization therapy implantable pacemaker	CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A25	No Apparent Adverse Event	02/08/2022	02/08/2022
							F26	No Health Consequences or Impact	A25	No Apparent Adverse Event	02/08/2022	02/08/2022
					CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A25	No Apparent Adverse Event	02/08/2022	02/08/2022
							F26	No Health Consequences or Impact	A25	No Apparent Adverse Event	02/08/2022	02/08/2022
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A25	No Apparent Adverse Event	02/08/2022	02/08/2022
							F26	No Health Consequences or Impact	A25	No Apparent Adverse Event	02/08/2022	02/08/2022
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A25	No Apparent Adverse Event	02/08/2022	02/08/2022

							F26	No Health Consequences or Impact	A25	No Apparent Adverse Event	02/08/2022	02/08/2022
13099	2023	All other reportable incidents	47205	Cardiac pulse generator programmer	CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A150105	Key or Button Unresponsive/not Working		
							F26	No Health Consequences or Impact	A150105	Key or Button Unresponsive/not Working		
13100	2023	All other reportable incidents	47205	Cardiac pulse generator programmer	CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	04/08/2022	04/08/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	04/08/2022	04/08/2022
13101	2023		47205	Cardiac pulse generator programmer	CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A090805	Incorrect Measurement		
							F23	Unexpected Medical Intervention	A090805	Incorrect Measurement		
13102	2023	All other reportable incidents	47263	Cardiac resynchronization therapy implantable pacemaker	CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	19/01/2023	19/01/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	19/01/2023	19/01/2023
					CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	19/01/2023	19/01/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	19/01/2023	19/01/2023
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	19/01/2023	19/01/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	19/01/2023	19/01/2023
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	19/01/2023	19/01/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	19/01/2023	19/01/2023
13103	2023	All other reportable incidents	47205	Cardiac pulse generator programmer	CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A090805	Incorrect Measurement		
							F23	Unexpected Medical Intervention	A090805	Incorrect Measurement		
13104	2023	All other reportable incidents	47263	Cardiac resynchronization therapy implantable pacemaker	CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	24/01/2023	24/01/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	24/01/2023	24/01/2023
					CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	24/01/2023	24/01/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	24/01/2023	24/01/2023
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	24/01/2023	24/01/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	24/01/2023	24/01/2023
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	24/01/2023	24/01/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	24/01/2023	24/01/2023
13105	2023	All other reportable incidents	47265	Dual-chamber implantable pacemaker, rate-responsive	CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem		
							F1901	Additional Surgery	A24	No Apparent Device Problem		
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem		
							F1901	Additional Surgery	A24	No Apparent Device Problem		
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem		
							F1901	Additional Surgery	A24	No Apparent Device Problem		
13106	2023	All other reportable incidents	47265	Dual-chamber implantable pacemaker, rate-responsive	CT141	Electrodes/Leads and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070909	Oversensing	30/09/2022	30/09/2022
									A090801	Signal Artifact/Noise	30/09/2022	30/09/2022
									A180104	Device Contamination with Chemical or Other Material	30/09/2022	30/09/2022
							F08	Hospitalization or Prolonged Hospitalization	A070909	Oversensing	30/09/2022	30/09/2022
									A090801	Signal Artifact/Noise	30/09/2022	30/09/2022
									A180104	Device Contamination with Chemical or Other Material	30/09/2022	30/09/2022
							F1905	Device Revision or Replacement	A070909	Oversensing	30/09/2022	30/09/2022
									A090801	Signal Artifact/Noise	30/09/2022	30/09/2022
									A180104	Device Contamination with Chemical or Other Material	30/09/2022	30/09/2022
							F2203	Imaging Required	A070909	Oversensing	30/09/2022	30/09/2022
									A090801	Signal Artifact/Noise	30/09/2022	30/09/2022
									A180104	Device Contamination with Chemical or Other Material	30/09/2022	30/09/2022
					CT145	Cardiovascular devices	E2403	No Clinical Signs, Symptoms or Conditions	A070909	Oversensing	30/09/2022	30/09/2022
									A090801	Signal Artifact/Noise	30/09/2022	30/09/2022
									A180104	Device Contamination with Chemical or Other Material	30/09/2022	30/09/2022
							F08	Hospitalization or Prolonged Hospitalization	A070909	Oversensing	30/09/2022	30/09/2022
									A090801	Signal Artifact/Noise	30/09/2022	30/09/2022
									A180104	Device Contamination with Chemical or Other Material	30/09/2022	30/09/2022
							F1905	Device Revision or Replacement	A070909	Oversensing	30/09/2022	30/09/2022
									A090801	Signal Artifact/Noise	30/09/2022	30/09/2022
									A180104	Device Contamination with Chemical or Other Material	30/09/2022	30/09/2022
							F2203	Imaging Required	A070909	Oversensing	30/09/2022	30/09/2022
									A090801	Signal Artifact/Noise	30/09/2022	30/09/2022
									A180104	Device Contamination with Chemical or Other Material	30/09/2022	30/09/2022
					CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070909	Oversensing	30/09/2022	30/09/2022
									A090801	Signal Artifact/Noise	30/09/2022	30/09/2022
									A180104	Device Contamination with Chemical or Other Material	30/09/2022	30/09/2022
							F08	Hospitalization or Prolonged Hospitalization	A070909	Oversensing	30/09/2022	30/09/2022
									A090801	Signal Artifact/Noise	30/09/2022	30/09/2022
									A180104	Device Contamination with Chemical or Other Material	30/09/2022	30/09/2022
							F1905	Device Revision or Replacement	A070909	Oversensing	30/09/2022	30/09/2022
									A090801	Signal Artifact/Noise	30/09/2022	30/09/2022
									A180104	Device Contamination with Chemical or Other Material	30/09/2022	30/09/2022
							F2203	Imaging Required	A070909	Oversensing	30/09/2022	30/09/2022
									A090801	Signal Artifact/Noise	30/09/2022	30/09/2022
									A180104	Device Contamination with Chemical or Other Material	30/09/2022	30/09/2022
					CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A070909	Oversensing	30/09/2022	30/09/2022
									A090801	Signal Artifact/Noise	30/09/2022	30/09/2022
									A180104	Device Contamination with Chemical or Other Material	30/09/2022	30/09/2022
							F08	Hospitalization or Prolonged Hospitalization	A070909	Oversensing	30/09/2022	30/09/2022
									A090801	Signal Artifact/Noise	30/09/2022	30/09/2022
									A180104	Device Contamination with Chemical or Other Material	30/09/2022	30/09/2022
							F1905	Device Revision or Replacement	A070909	Oversensing	30/09/2022	30/09/2022
									A090801	Signal Artifact/Noise	30/09/2022	30/09/2022
									A180104	Device Contamination with Chemical or Other Material	30/09/2022	30/09/2022

							F2203	Imaging Required	A070909	Oversensing	30/09/2022	30/09/2022
									A090801	Signal Artifact/Noise	30/09/2022	30/09/2022
									A180104	Device Contamination with Chemical or Other Material	30/09/2022	30/09/2022
					CT233	Surgical	E2403	No Clinical Signs, Symptoms or Conditions	A070909	Oversensing	30/09/2022	30/09/2022
									A090801	Signal Artifact/Noise	30/09/2022	30/09/2022
									A180104	Device Contamination with Chemical or Other Material	30/09/2022	30/09/2022
							F08	Hospitalization or Prolonged Hospitalization	A070909	Oversensing	30/09/2022	30/09/2022
									A090801	Signal Artifact/Noise	30/09/2022	30/09/2022
									A180104	Device Contamination with Chemical or Other Material	30/09/2022	30/09/2022
							F1905	Device Revision or Replacement	A070909	Oversensing	30/09/2022	30/09/2022
									A090801	Signal Artifact/Noise	30/09/2022	30/09/2022
									A180104	Device Contamination with Chemical or Other Material	30/09/2022	30/09/2022
							F2203	Imaging Required	A070909	Oversensing	30/09/2022	30/09/2022
									A090801	Signal Artifact/Noise	30/09/2022	30/09/2022
									A180104	Device Contamination with Chemical or Other Material	30/09/2022	30/09/2022
					CT2406	Implantable	E2403	No Clinical Signs, Symptoms or Conditions	A070909	Oversensing	30/09/2022	30/09/2022
									A090801	Signal Artifact/Noise	30/09/2022	30/09/2022
									A180104	Device Contamination with Chemical or Other Material	30/09/2022	30/09/2022
							F08	Hospitalization or Prolonged Hospitalization	A070909	Oversensing	30/09/2022	30/09/2022
									A090801	Signal Artifact/Noise	30/09/2022	30/09/2022
									A180104	Device Contamination with Chemical or Other Material	30/09/2022	30/09/2022
							F1905	Device Revision or Replacement	A070909	Oversensing	30/09/2022	30/09/2022
									A090801	Signal Artifact/Noise	30/09/2022	30/09/2022
									A180104	Device Contamination with Chemical or Other Material	30/09/2022	30/09/2022
							F2203	Imaging Required	A070909	Oversensing	30/09/2022	30/09/2022
									A090801	Signal Artifact/Noise	30/09/2022	30/09/2022
									A180104	Device Contamination with Chemical or Other Material	30/09/2022	30/09/2022
					CT2407	Active implantable	E2403	No Clinical Signs, Symptoms or Conditions	A070909	Oversensing	30/09/2022	30/09/2022
									A090801	Signal Artifact/Noise	30/09/2022	30/09/2022
									A180104	Device Contamination with Chemical or Other Material	30/09/2022	30/09/2022
							F08	Hospitalization or Prolonged Hospitalization	A070909	Oversensing	30/09/2022	30/09/2022
									A090801	Signal Artifact/Noise	30/09/2022	30/09/2022
									A180104	Device Contamination with Chemical or Other Material	30/09/2022	30/09/2022
							F1905	Device Revision or Replacement	A070909	Oversensing	30/09/2022	30/09/2022
									A090801	Signal Artifact/Noise	30/09/2022	30/09/2022
									A180104	Device Contamination with Chemical or Other Material	30/09/2022	30/09/2022
							F2203	Imaging Required	A070909	Oversensing	30/09/2022	30/09/2022
									A090801	Signal Artifact/Noise	30/09/2022	30/09/2022
									A180104	Device Contamination with Chemical or Other Material	30/09/2022	30/09/2022
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A070909	Oversensing	30/09/2022	30/09/2022
									A090801	Signal Artifact/Noise	30/09/2022	30/09/2022
									A180104	Device Contamination with Chemical or Other Material	30/09/2022	30/09/2022
							F08	Hospitalization or Prolonged Hospitalization	A070909	Oversensing	30/09/2022	30/09/2022
									A090801	Signal Artifact/Noise	30/09/2022	30/09/2022
									A180104	Device Contamination with Chemical or Other Material	30/09/2022	30/09/2022
							F1905	Device Revision or Replacement	A070909	Oversensing	30/09/2022	30/09/2022
									A090801	Signal Artifact/Noise	30/09/2022	30/09/2022
									A180104	Device Contamination with Chemical or Other Material	30/09/2022	30/09/2022
							F2203	Imaging Required	A070909	Oversensing	30/09/2022	30/09/2022
									A090801	Signal Artifact/Noise	30/09/2022	30/09/2022
									A180104	Device Contamination with Chemical or Other Material	30/09/2022	30/09/2022
					CT2727	Active implant devices/ancillaries	E2403	No Clinical Signs, Symptoms or Conditions	A070909	Oversensing	30/09/2022	30/09/2022
									A090801	Signal Artifact/Noise	30/09/2022	30/09/2022
									A180104	Device Contamination with Chemical or Other Material	30/09/2022	30/09/2022
							F08	Hospitalization or Prolonged Hospitalization	A070909	Oversensing	30/09/2022	30/09/2022
									A090801	Signal Artifact/Noise	30/09/2022	30/09/2022
									A180104	Device Contamination with Chemical or Other Material	30/09/2022	30/09/2022
							F1905	Device Revision or Replacement	A070909	Oversensing	30/09/2022	30/09/2022
									A090801	Signal Artifact/Noise	30/09/2022	30/09/2022
									A180104	Device Contamination with Chemical or Other Material	30/09/2022	30/09/2022
							F2203	Imaging Required	A070909	Oversensing	30/09/2022	30/09/2022
									A090801	Signal Artifact/Noise	30/09/2022	30/09/2022
									A180104	Device Contamination with Chemical or Other Material	30/09/2022	30/09/2022
					CT2750	Implantable cardiac electrical stimulators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070909	Oversensing	30/09/2022	30/09/2022
									A090801	Signal Artifact/Noise	30/09/2022	30/09/2022
									A180104	Device Contamination with Chemical or Other Material	30/09/2022	30/09/2022
							F08	Hospitalization or Prolonged Hospitalization	A070909	Oversensing	30/09/2022	30/09/2022
									A090801	Signal Artifact/Noise	30/09/2022	30/09/2022
									A180104	Device Contamination with Chemical or Other Material	30/09/2022	30/09/2022
							F1905	Device Revision or Replacement	A070909	Oversensing	30/09/2022	30/09/2022

								A090801	Signal Artifact/Noise	30/09/2022	30/09/2022
								A180104	Device Contamination with Chemical or Other Material	30/09/2022	30/09/2022
						F2203	Imaging Required	A070909	Oversensing	30/09/2022	30/09/2022
								A090801	Signal Artifact/Noise	30/09/2022	30/09/2022
								A180104	Device Contamination with Chemical or Other Material	30/09/2022	30/09/2022
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A070909	Oversensing	30/09/2022	30/09/2022
									A090801	Signal Artifact/Noise	30/09/2022	30/09/2022
									A180104	Device Contamination with Chemical or Other Material	30/09/2022	30/09/2022
							F08	Hospitalization or Prolonged Hospitalization	A070909	Oversensing	30/09/2022	30/09/2022
									A090801	Signal Artifact/Noise	30/09/2022	30/09/2022
									A180104	Device Contamination with Chemical or Other Material	30/09/2022	30/09/2022
							F1905	Device Revision or Replacement	A070909	Oversensing	30/09/2022	30/09/2022
									A090801	Signal Artifact/Noise	30/09/2022	30/09/2022
									A180104	Device Contamination with Chemical or Other Material	30/09/2022	30/09/2022
							F2203	Imaging Required	A070909	Oversensing	30/09/2022	30/09/2022
									A090801	Signal Artifact/Noise	30/09/2022	30/09/2022
									A180104	Device Contamination with Chemical or Other Material	30/09/2022	30/09/2022
					CT2985	Active	E2403	No Clinical Signs, Symptoms or Conditions	A070909	Oversensing	30/09/2022	30/09/2022
									A090801	Signal Artifact/Noise	30/09/2022	30/09/2022
									A180104	Device Contamination with Chemical or Other Material	30/09/2022	30/09/2022
							F08	Hospitalization or Prolonged Hospitalization	A070909	Oversensing	30/09/2022	30/09/2022
									A090801	Signal Artifact/Noise	30/09/2022	30/09/2022
									A180104	Device Contamination with Chemical or Other Material	30/09/2022	30/09/2022
							F1905	Device Revision or Replacement	A070909	Oversensing	30/09/2022	30/09/2022
									A090801	Signal Artifact/Noise	30/09/2022	30/09/2022
									A180104	Device Contamination with Chemical or Other Material	30/09/2022	30/09/2022
							F2203	Imaging Required	A070909	Oversensing	30/09/2022	30/09/2022
									A090801	Signal Artifact/Noise	30/09/2022	30/09/2022
									A180104	Device Contamination with Chemical or Other Material	30/09/2022	30/09/2022
					CT3001	Implantable metal	E2403	No Clinical Signs, Symptoms or Conditions	A070909	Oversensing	30/09/2022	30/09/2022
									A090801	Signal Artifact/Noise	30/09/2022	30/09/2022
									A180104	Device Contamination with Chemical or Other Material	30/09/2022	30/09/2022
							F08	Hospitalization or Prolonged Hospitalization	A070909	Oversensing	30/09/2022	30/09/2022
									A090801	Signal Artifact/Noise	30/09/2022	30/09/2022
									A180104	Device Contamination with Chemical or Other Material	30/09/2022	30/09/2022
							F1905	Device Revision or Replacement	A070909	Oversensing	30/09/2022	30/09/2022
									A090801	Signal Artifact/Noise	30/09/2022	30/09/2022
									A180104	Device Contamination with Chemical or Other Material	30/09/2022	30/09/2022
							F2203	Imaging Required	A070909	Oversensing	30/09/2022	30/09/2022
									A090801	Signal Artifact/Noise	30/09/2022	30/09/2022
									A180104	Device Contamination with Chemical or Other Material	30/09/2022	30/09/2022
					CT3002	Implantable synthetic polymer	E2403	No Clinical Signs, Symptoms or Conditions	A070909	Oversensing	30/09/2022	30/09/2022
									A090801	Signal Artifact/Noise	30/09/2022	30/09/2022
									A180104	Device Contamination with Chemical or Other Material	30/09/2022	30/09/2022
							F08	Hospitalization or Prolonged Hospitalization	A070909	Oversensing	30/09/2022	30/09/2022
									A090801	Signal Artifact/Noise	30/09/2022	30/09/2022
									A180104	Device Contamination with Chemical or Other Material	30/09/2022	30/09/2022
							F1905	Device Revision or Replacement	A070909	Oversensing	30/09/2022	30/09/2022
									A090801	Signal Artifact/Noise	30/09/2022	30/09/2022
									A180104	Device Contamination with Chemical or Other Material	30/09/2022	30/09/2022
							F2203	Imaging Required	A070909	Oversensing	30/09/2022	30/09/2022
									A090801	Signal Artifact/Noise	30/09/2022	30/09/2022
									A180104	Device Contamination with Chemical or Other Material	30/09/2022	30/09/2022
					CT3006	Implantable inorganic material	E2403	No Clinical Signs, Symptoms or Conditions	A070909	Oversensing	30/09/2022	30/09/2022
									A090801	Signal Artifact/Noise	30/09/2022	30/09/2022
									A180104	Device Contamination with Chemical or Other Material	30/09/2022	30/09/2022
							F08	Hospitalization or Prolonged Hospitalization	A070909	Oversensing	30/09/2022	30/09/2022
									A090801	Signal Artifact/Noise	30/09/2022	30/09/2022
									A180104	Device Contamination with Chemical or Other Material	30/09/2022	30/09/2022
							F1905	Device Revision or Replacement	A070909	Oversensing	30/09/2022	30/09/2022
									A090801	Signal Artifact/Noise	30/09/2022	30/09/2022
									A180104	Device Contamination with Chemical or Other Material	30/09/2022	30/09/2022
							F2203	Imaging Required	A070909	Oversensing	30/09/2022	30/09/2022
									A090801	Signal Artifact/Noise	30/09/2022	30/09/2022
									A180104	Device Contamination with Chemical or Other Material	30/09/2022	30/09/2022
					CT314	Electrically powered	E2403	No Clinical Signs, Symptoms or Conditions	A070909	Oversensing	30/09/2022	30/09/2022
									A090801	Signal Artifact/Noise	30/09/2022	30/09/2022
									A180104	Device Contamination with Chemical or Other Material	30/09/2022	30/09/2022
							F08	Hospitalization or Prolonged Hospitalization	A070909	Oversensing	30/09/2022	30/09/2022
									A090801	Signal Artifact/Noise	30/09/2022	30/09/2022

									A180104	Device Contamination with Chemical or Other Material	30/09/2022	30/09/2022
							F1905	Device Revision or Replacement	A070909	Oversensing	30/09/2022	30/09/2022
									A090801	Signal Artifact/Noise	30/09/2022	30/09/2022
									A180104	Device Contamination with Chemical or Other Material	30/09/2022	30/09/2022
							F2203	Imaging Required	A070909	Oversensing	30/09/2022	30/09/2022
									A090801	Signal Artifact/Noise	30/09/2022	30/09/2022
									A180104	Device Contamination with Chemical or Other Material	30/09/2022	30/09/2022
					CT321	Long-term surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A070909	Oversensing	30/09/2022	30/09/2022
									A090801	Signal Artifact/Noise	30/09/2022	30/09/2022
									A180104	Device Contamination with Chemical or Other Material	30/09/2022	30/09/2022
							F08	Hospitalization or Prolonged Hospitalization	A070909	Oversensing	30/09/2022	30/09/2022
									A090801	Signal Artifact/Noise	30/09/2022	30/09/2022
									A180104	Device Contamination with Chemical or Other Material	30/09/2022	30/09/2022
							F1905	Device Revision or Replacement	A070909	Oversensing	30/09/2022	30/09/2022
									A090801	Signal Artifact/Noise	30/09/2022	30/09/2022
									A180104	Device Contamination with Chemical or Other Material	30/09/2022	30/09/2022
							F2203	Imaging Required	A070909	Oversensing	30/09/2022	30/09/2022
									A090801	Signal Artifact/Noise	30/09/2022	30/09/2022
									A180104	Device Contamination with Chemical or Other Material	30/09/2022	30/09/2022
					CT449	Generators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070909	Oversensing	30/09/2022	30/09/2022
									A090801	Signal Artifact/Noise	30/09/2022	30/09/2022
									A180104	Device Contamination with Chemical or Other Material	30/09/2022	30/09/2022
							F08	Hospitalization or Prolonged Hospitalization	A070909	Oversensing	30/09/2022	30/09/2022
									A090801	Signal Artifact/Noise	30/09/2022	30/09/2022
									A180104	Device Contamination with Chemical or Other Material	30/09/2022	30/09/2022
							F1905	Device Revision or Replacement	A070909	Oversensing	30/09/2022	30/09/2022
									A090801	Signal Artifact/Noise	30/09/2022	30/09/2022
									A180104	Device Contamination with Chemical or Other Material	30/09/2022	30/09/2022
							F2203	Imaging Required	A070909	Oversensing	30/09/2022	30/09/2022
									A090801	Signal Artifact/Noise	30/09/2022	30/09/2022
									A180104	Device Contamination with Chemical or Other Material	30/09/2022	30/09/2022
					CT981	Single use	E2403	No Clinical Signs, Symptoms or Conditions	A070909	Oversensing	30/09/2022	30/09/2022
									A090801	Signal Artifact/Noise	30/09/2022	30/09/2022
									A180104	Device Contamination with Chemical or Other Material	30/09/2022	30/09/2022
							F08	Hospitalization or Prolonged Hospitalization	A070909	Oversensing	30/09/2022	30/09/2022
									A090801	Signal Artifact/Noise	30/09/2022	30/09/2022
									A180104	Device Contamination with Chemical or Other Material	30/09/2022	30/09/2022
							F1905	Device Revision or Replacement	A070909	Oversensing	30/09/2022	30/09/2022
									A090801	Signal Artifact/Noise	30/09/2022	30/09/2022
									A180104	Device Contamination with Chemical or Other Material	30/09/2022	30/09/2022
							F2203	Imaging Required	A070909	Oversensing	30/09/2022	30/09/2022
									A090801	Signal Artifact/Noise	30/09/2022	30/09/2022
									A180104	Device Contamination with Chemical or Other Material	30/09/2022	30/09/2022
					CT983	Surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A070909	Oversensing	30/09/2022	30/09/2022
									A090801	Signal Artifact/Noise	30/09/2022	30/09/2022
									A180104	Device Contamination with Chemical or Other Material	30/09/2022	30/09/2022
							F08	Hospitalization or Prolonged Hospitalization	A070909	Oversensing	30/09/2022	30/09/2022
									A090801	Signal Artifact/Noise	30/09/2022	30/09/2022
									A180104	Device Contamination with Chemical or Other Material	30/09/2022	30/09/2022
							F1905	Device Revision or Replacement	A070909	Oversensing	30/09/2022	30/09/2022
									A090801	Signal Artifact/Noise	30/09/2022	30/09/2022
									A180104	Device Contamination with Chemical or Other Material	30/09/2022	30/09/2022
							F2203	Imaging Required	A070909	Oversensing	30/09/2022	30/09/2022
									A090801	Signal Artifact/Noise	30/09/2022	30/09/2022
									A180104	Device Contamination with Chemical or Other Material	30/09/2022	30/09/2022
13107	2023	All other reportable incidents	47263	Cardiac resynchronization therapy implantable pacemaker	CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A0908	Incorrect, Inadequate or Imprecise Result or Readings		
							F26	No Health Consequences or Impact	A0908	Incorrect, Inadequate or Imprecise Result or Readings		
					CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A0908	Incorrect, Inadequate or Imprecise Result or Readings		
							F26	No Health Consequences or Impact	A0908	Incorrect, Inadequate or Imprecise Result or Readings		
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A0908	Incorrect, Inadequate or Imprecise Result or Readings		
							F26	No Health Consequences or Impact	A0908	Incorrect, Inadequate or Imprecise Result or Readings		
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A0908	Incorrect, Inadequate or Imprecise Result or Readings		
							F26	No Health Consequences or Impact	A0908	Incorrect, Inadequate or Imprecise Result or Readings		
13108	2023	All other reportable incidents	35852	Single-chamber implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A070909	Oversensing		
									A0712	Pacing Problem		
							F26	No Health Consequences or Impact	A070909	Oversensing		
									A0712	Pacing Problem		

13109	2023	All other reportable incidents	35852	Single-chamber implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
									A0908	Incorrect, Inadequate or Imprecise Result or Readings		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
									A0908	Incorrect, Inadequate or Imprecise Result or Readings		
13110	2023	All other reportable incidents	47265	Dual-chamber implantable pacemaker, rate-responsive	CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
13111	2023		47270	Cardiac resynchronization therapy implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E2107	Failure of Implant	A0203	Defective Device		
									A07	Electrical /Electronic Property Problem		
									A1910	Electrical Power Problem		
							F07	Exacerbation of Existing Condition	A0203	Defective Device		
									A07	Electrical /Electronic Property Problem		
									A1910	Electrical Power Problem		
							F15	Recognized Device or Procedural Complication	A0203	Defective Device		
									A07	Electrical /Electronic Property Problem		
									A1910	Electrical Power Problem		
							F16	Reduction in Life Expectancy	A0203	Defective Device		
									A07	Electrical /Electronic Property Problem		
									A1910	Electrical Power Problem		
13112	2023	All other reportable incidents	35223	Endocardial/interventricular septal pacing lead	CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A040101	Fracture	18/04/2023	18/04/2023
									A070909	Oversensing	18/04/2023	18/04/2023
							F1901	Additional Surgery	A040101	Fracture	18/04/2023	18/04/2023
									A070909	Oversensing	18/04/2023	18/04/2023
13113	2023	Death	60789	Intracardiac pacemaker	CT130	Catheters and associated devices	E0109	Convulsion/Seizure	A24	No Apparent Device Problem		
							E0306	Sepsis	A24	No Apparent Device Problem		
							E0602	Cardiac Arrest	A24	No Apparent Device Problem		
							E0619	Pericardial Effusion	A24	No Apparent Device Problem		
							E0743	Respiratory Insufficiency	A24	No Apparent Device Problem		
							E233605	Septic Shock	A24	No Apparent Device Problem		
							F02	Death	A24	No Apparent Device Problem		
							F23	Unexpected Medical Intervention	A24	No Apparent Device Problem		
							F2303	Medication Required	A24	No Apparent Device Problem		
					CT145	Cardiovascular devices	E0109	Convulsion/Seizure	A24	No Apparent Device Problem		
							E0306	Sepsis	A24	No Apparent Device Problem		
							E0602	Cardiac Arrest	A24	No Apparent Device Problem		
							E0619	Pericardial Effusion	A24	No Apparent Device Problem		
							E0743	Respiratory Insufficiency	A24	No Apparent Device Problem		
							E233605	Septic Shock	A24	No Apparent Device Problem		
							F02	Death	A24	No Apparent Device Problem		
							F23	Unexpected Medical Intervention	A24	No Apparent Device Problem		
							F2303	Medication Required	A24	No Apparent Device Problem		
					CT1684	Cardiac pacemakers and associated devices	E0109	Convulsion/Seizure	A24	No Apparent Device Problem		
							E0306	Sepsis	A24	No Apparent Device Problem		
							E0602	Cardiac Arrest	A24	No Apparent Device Problem		
							E0619	Pericardial Effusion	A24	No Apparent Device Problem		
							E0743	Respiratory Insufficiency	A24	No Apparent Device Problem		
							E233605	Septic Shock	A24	No Apparent Device Problem		
							F02	Death	A24	No Apparent Device Problem		
							F23	Unexpected Medical Intervention	A24	No Apparent Device Problem		
							F2303	Medication Required	A24	No Apparent Device Problem		
					CT1685	Cardiac pacemakers	E0109	Convulsion/Seizure	A24	No Apparent Device Problem		
							E0306	Sepsis	A24	No Apparent Device Problem		
							E0602	Cardiac Arrest	A24	No Apparent Device Problem		
							E0619	Pericardial Effusion	A24	No Apparent Device Problem		
							E0743	Respiratory Insufficiency	A24	No Apparent Device Problem		
							E233605	Septic Shock	A24	No Apparent Device Problem		
							F02	Death	A24	No Apparent Device Problem		
							F23	Unexpected Medical Intervention	A24	No Apparent Device Problem		
							F2303	Medication Required	A24	No Apparent Device Problem		
					CT233	Surgical	E0109	Convulsion/Seizure	A24	No Apparent Device Problem		
							E0306	Sepsis	A24	No Apparent Device Problem		
							E0602	Cardiac Arrest	A24	No Apparent Device Problem		
							E0619	Pericardial Effusion	A24	No Apparent Device Problem		
							E0743	Respiratory Insufficiency	A24	No Apparent Device Problem		
							E233605	Septic Shock	A24	No Apparent Device Problem		
							F02	Death	A24	No Apparent Device Problem		
							F23	Unexpected Medical Intervention	A24	No Apparent Device Problem		
							F2303	Medication Required	A24	No Apparent Device Problem		
					CT2406	Implantable	E0109	Convulsion/Seizure	A24	No Apparent Device Problem		
							E0306	Sepsis	A24	No Apparent Device Problem		
							E0602	Cardiac Arrest	A24	No Apparent Device Problem		
							E0619	Pericardial Effusion	A24	No Apparent Device Problem		
							E0743	Respiratory Insufficiency	A24	No Apparent Device Problem		
							E233605	Septic Shock	A24	No Apparent Device Problem		
							F02	Death	A24	No Apparent Device Problem		
							F23	Unexpected Medical Intervention	A24	No Apparent Device Problem		
							F2303	Medication Required	A24	No Apparent Device Problem		

					CT2407	Active implantable	E0109	Convulsion/Seizure	A24	No Apparent Device Problem		
							E0306	Sepsis	A24	No Apparent Device Problem		
							E0602	Cardiac Arrest	A24	No Apparent Device Problem		
							E0619	Pericardial Effusion	A24	No Apparent Device Problem		
							E0743	Respiratory Insufficiency	A24	No Apparent Device Problem		
							E233605	Septic Shock	A24	No Apparent Device Problem		
							F02	Death	A24	No Apparent Device Problem		
							F23	Unexpected Medical Intervention	A24	No Apparent Device Problem		
							F2303	Medication Required	A24	No Apparent Device Problem		
					CT257	Implantable cardiac pacemakers	E0109	Convulsion/Seizure	A24	No Apparent Device Problem		
							E0306	Sepsis	A24	No Apparent Device Problem		
							E0602	Cardiac Arrest	A24	No Apparent Device Problem		
							E0619	Pericardial Effusion	A24	No Apparent Device Problem		
							E0743	Respiratory Insufficiency	A24	No Apparent Device Problem		
							E233605	Septic Shock	A24	No Apparent Device Problem		
							F02	Death	A24	No Apparent Device Problem		
							F23	Unexpected Medical Intervention	A24	No Apparent Device Problem		
							F2303	Medication Required	A24	No Apparent Device Problem		
					CT2727	Active implant devices/ancillaries	E0109	Convulsion/Seizure	A24	No Apparent Device Problem		
							E0306	Sepsis	A24	No Apparent Device Problem		
							E0602	Cardiac Arrest	A24	No Apparent Device Problem		
							E0619	Pericardial Effusion	A24	No Apparent Device Problem		
							E0743	Respiratory Insufficiency	A24	No Apparent Device Problem		
							E233605	Septic Shock	A24	No Apparent Device Problem		
							F02	Death	A24	No Apparent Device Problem		
							F23	Unexpected Medical Intervention	A24	No Apparent Device Problem		
							F2303	Medication Required	A24	No Apparent Device Problem		
					CT2750	Implantable cardiac electrical stimulators and associated devices	E0109	Convulsion/Seizure	A24	No Apparent Device Problem		
							E0306	Sepsis	A24	No Apparent Device Problem		
							E0602	Cardiac Arrest	A24	No Apparent Device Problem		
							E0619	Pericardial Effusion	A24	No Apparent Device Problem		
							E0743	Respiratory Insufficiency	A24	No Apparent Device Problem		
							E233605	Septic Shock	A24	No Apparent Device Problem		
							F02	Death	A24	No Apparent Device Problem		
							F23	Unexpected Medical Intervention	A24	No Apparent Device Problem		
							F2303	Medication Required	A24	No Apparent Device Problem		
					CT2818	Catheterization/Cannulation/Fluid path devices	E0109	Convulsion/Seizure	A24	No Apparent Device Problem		
							E0306	Sepsis	A24	No Apparent Device Problem		
							E0602	Cardiac Arrest	A24	No Apparent Device Problem		
							E0619	Pericardial Effusion	A24	No Apparent Device Problem		
							E0743	Respiratory Insufficiency	A24	No Apparent Device Problem		
							E233605	Septic Shock	A24	No Apparent Device Problem		
							F02	Death	A24	No Apparent Device Problem		
							F23	Unexpected Medical Intervention	A24	No Apparent Device Problem		
							F2303	Medication Required	A24	No Apparent Device Problem		
					CT2866	Implantable cardiac electrical stimulators	E0109	Convulsion/Seizure	A24	No Apparent Device Problem		
							E0306	Sepsis	A24	No Apparent Device Problem		
							E0602	Cardiac Arrest	A24	No Apparent Device Problem		
							E0619	Pericardial Effusion	A24	No Apparent Device Problem		
							E0743	Respiratory Insufficiency	A24	No Apparent Device Problem		
							E233605	Septic Shock	A24	No Apparent Device Problem		
							F02	Death	A24	No Apparent Device Problem		
							F23	Unexpected Medical Intervention	A24	No Apparent Device Problem		
							F2303	Medication Required	A24	No Apparent Device Problem		
					CT2985	Active	E0109	Convulsion/Seizure	A24	No Apparent Device Problem		
							E0306	Sepsis	A24	No Apparent Device Problem		
							E0602	Cardiac Arrest	A24	No Apparent Device Problem		
							E0619	Pericardial Effusion	A24	No Apparent Device Problem		
							E0743	Respiratory Insufficiency	A24	No Apparent Device Problem		
							E233605	Septic Shock	A24	No Apparent Device Problem		
							F02	Death	A24	No Apparent Device Problem		
							F23	Unexpected Medical Intervention	A24	No Apparent Device Problem		
							F2303	Medication Required	A24	No Apparent Device Problem		
					CT3001	Implantable metal	E0109	Convulsion/Seizure	A24	No Apparent Device Problem		
							E0306	Sepsis	A24	No Apparent Device Problem		
							E0602	Cardiac Arrest	A24	No Apparent Device Problem		
							E0619	Pericardial Effusion	A24	No Apparent Device Problem		
							E0743	Respiratory Insufficiency	A24	No Apparent Device Problem		
							E233605	Septic Shock	A24	No Apparent Device Problem		
							F02	Death	A24	No Apparent Device Problem		
							F23	Unexpected Medical Intervention	A24	No Apparent Device Problem		
							F2303	Medication Required	A24	No Apparent Device Problem		
					CT3002	Implantable synthetic polymer	E0109	Convulsion/Seizure	A24	No Apparent Device Problem		
							E0306	Sepsis	A24	No Apparent Device Problem		
							E0602	Cardiac Arrest	A24	No Apparent Device Problem		
							E0619	Pericardial Effusion	A24	No Apparent Device Problem		
							E0743	Respiratory Insufficiency	A24	No Apparent Device Problem		
							E233605	Septic Shock	A24	No Apparent Device Problem		
							F02	Death	A24	No Apparent Device Problem		
							F23	Unexpected Medical Intervention	A24	No Apparent Device Problem		
							F2303	Medication Required	A24	No Apparent Device Problem		
					CT3006	Implantable inorganic material	E0109	Convulsion/Seizure	A24	No Apparent Device Problem		
							E0306	Sepsis	A24	No Apparent Device Problem		
							E0602	Cardiac Arrest	A24	No Apparent Device Problem		
							E0619	Pericardial Effusion	A24	No Apparent Device Problem		
							E0743	Respiratory Insufficiency	A24	No Apparent Device Problem		
							E233605	Septic Shock	A24	No Apparent Device Problem		
							F02	Death	A24	No Apparent Device Problem		

							F23	Unexpected Medical Intervention	A24	No Apparent Device Problem		
							F2303	Medication Required	A24	No Apparent Device Problem		
					CT314	Electrically powered	E0109	Convulsion/Seizure	A24	No Apparent Device Problem		
							E0306	Sepsis	A24	No Apparent Device Problem		
							E0602	Cardiac Arrest	A24	No Apparent Device Problem		
							E0619	Pericardial Effusion	A24	No Apparent Device Problem		
							E0743	Respiratory Insufficiency	A24	No Apparent Device Problem		
							E233605	Septic Shock	A24	No Apparent Device Problem		
							F02	Death	A24	No Apparent Device Problem		
							F23	Unexpected Medical Intervention	A24	No Apparent Device Problem		
							F2303	Medication Required	A24	No Apparent Device Problem		
					CT321	Long-term surgical invasive	E0109	Convulsion/Seizure	A24	No Apparent Device Problem		
							E0306	Sepsis	A24	No Apparent Device Problem		
							E0602	Cardiac Arrest	A24	No Apparent Device Problem		
							E0619	Pericardial Effusion	A24	No Apparent Device Problem		
							E0743	Respiratory Insufficiency	A24	No Apparent Device Problem		
							E233605	Septic Shock	A24	No Apparent Device Problem		
							F02	Death	A24	No Apparent Device Problem		
							F23	Unexpected Medical Intervention	A24	No Apparent Device Problem		
							F2303	Medication Required	A24	No Apparent Device Problem		
					CT981	Single use	E0109	Convulsion/Seizure	A24	No Apparent Device Problem		
							E0306	Sepsis	A24	No Apparent Device Problem		
							E0602	Cardiac Arrest	A24	No Apparent Device Problem		
							E0619	Pericardial Effusion	A24	No Apparent Device Problem		
							E0743	Respiratory Insufficiency	A24	No Apparent Device Problem		
							E233605	Septic Shock	A24	No Apparent Device Problem		
							F02	Death	A24	No Apparent Device Problem		
							F23	Unexpected Medical Intervention	A24	No Apparent Device Problem		
							F2303	Medication Required	A24	No Apparent Device Problem		
					CT983	Surgical invasive	E0109	Convulsion/Seizure	A24	No Apparent Device Problem		
							E0306	Sepsis	A24	No Apparent Device Problem		
							E0602	Cardiac Arrest	A24	No Apparent Device Problem		
							E0619	Pericardial Effusion	A24	No Apparent Device Problem		
							E0743	Respiratory Insufficiency	A24	No Apparent Device Problem		
							E233605	Septic Shock	A24	No Apparent Device Problem		
							F02	Death	A24	No Apparent Device Problem		
							F23	Unexpected Medical Intervention	A24	No Apparent Device Problem		
							F2303	Medication Required	A24	No Apparent Device Problem		
13114	2023	All other reportable incidents	35852	Single-chamber implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A07050	Battery Problem		
									A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A07050	Battery Problem		
									A070504	Premature Discharge of Battery		
13115	2023	Death	47265	Dual-chamber implantable pacemaker, rate-responsive	CT141	Electrodes/Leads and associated devices	E060110	Ventricular Fibrillation	A24	No Apparent Device Problem		
							E0602	Cardiac Arrest	A24	No Apparent Device Problem		
							F02	Death	A24	No Apparent Device Problem		
					CT145	Cardiovascular devices	E060110	Ventricular Fibrillation	A24	No Apparent Device Problem		
							E0602	Cardiac Arrest	A24	No Apparent Device Problem		
							F02	Death	A24	No Apparent Device Problem		
					CT1684	Cardiac pacemakers and associated devices	E060110	Ventricular Fibrillation	A24	No Apparent Device Problem		
							E0602	Cardiac Arrest	A24	No Apparent Device Problem		
							F02	Death	A24	No Apparent Device Problem		
					CT1685	Cardiac pacemakers	E060110	Ventricular Fibrillation	A24	No Apparent Device Problem		
							E0602	Cardiac Arrest	A24	No Apparent Device Problem		
							F02	Death	A24	No Apparent Device Problem		
					CT233	Surgical	E060110	Ventricular Fibrillation	A24	No Apparent Device Problem		
							E0602	Cardiac Arrest	A24	No Apparent Device Problem		
							F02	Death	A24	No Apparent Device Problem		
					CT2406	Implantable	E060110	Ventricular Fibrillation	A24	No Apparent Device Problem		
							E0602	Cardiac Arrest	A24	No Apparent Device Problem		
							F02	Death	A24	No Apparent Device Problem		
					CT2407	Active implantable	E060110	Ventricular Fibrillation	A24	No Apparent Device Problem		
							E0602	Cardiac Arrest	A24	No Apparent Device Problem		
							F02	Death	A24	No Apparent Device Problem		
					CT257	Implantable cardiac pacemakers	E060110	Ventricular Fibrillation	A24	No Apparent Device Problem		
							E0602	Cardiac Arrest	A24	No Apparent Device Problem		
							F02	Death	A24	No Apparent Device Problem		
					CT2727	Active implant devices/ancillaries	E060110	Ventricular Fibrillation	A24	No Apparent Device Problem		
							E0602	Cardiac Arrest	A24	No Apparent Device Problem		
							F02	Death	A24	No Apparent Device Problem		
					CT2750	Implantable cardiac electrical stimulators and associated devices	E060110	Ventricular Fibrillation	A24	No Apparent Device Problem		
							E0602	Cardiac Arrest	A24	No Apparent Device Problem		
							F02	Death	A24	No Apparent Device Problem		
					CT2866	Implantable cardiac electrical stimulators	E060110	Ventricular Fibrillation	A24	No Apparent Device Problem		
							E0602	Cardiac Arrest	A24	No Apparent Device Problem		
							F02	Death	A24	No Apparent Device Problem		
					CT2985	Active	E060110	Ventricular Fibrillation	A24	No Apparent Device Problem		
							E0602	Cardiac Arrest	A24	No Apparent Device Problem		
							F02	Death	A24	No Apparent Device Problem		
					CT3001	Implantable metal	E060110	Ventricular Fibrillation	A24	No Apparent Device Problem		
							E0602	Cardiac Arrest	A24	No Apparent Device Problem		
							F02	Death	A24	No Apparent Device Problem		
					CT3002	Implantable synthetic polymer	E060110	Ventricular Fibrillation	A24	No Apparent Device Problem		
							E0602	Cardiac Arrest	A24	No Apparent Device Problem		
							F02	Death	A24	No Apparent Device Problem		
					CT3006	Implantable inorganic material	E060110	Ventricular Fibrillation	A24	No Apparent Device Problem		
							E0602	Cardiac Arrest	A24	No Apparent Device Problem		
							F02	Death	A24	No Apparent Device Problem		
					CT314	Electrically powered	E060110	Ventricular Fibrillation	A24	No Apparent Device Problem		

							E0602	Cardiac Arrest	A24	No Apparent Device Problem		
							F02	Death	A24	No Apparent Device Problem		
					CT321	Long-term surgical invasive	E060110	Ventricular Fibrillation	A24	No Apparent Device Problem		
							E0602	Cardiac Arrest	A24	No Apparent Device Problem		
							F02	Death	A24	No Apparent Device Problem		
					CT449	Generators and associated devices	E060110	Ventricular Fibrillation	A24	No Apparent Device Problem		
							E0602	Cardiac Arrest	A24	No Apparent Device Problem		
							F02	Death	A24	No Apparent Device Problem		
					CT981	Single use	E060110	Ventricular Fibrillation	A24	No Apparent Device Problem		
							E0602	Cardiac Arrest	A24	No Apparent Device Problem		
							F02	Death	A24	No Apparent Device Problem		
					CT983	Surgical invasive	E060110	Ventricular Fibrillation	A24	No Apparent Device Problem		
							E0602	Cardiac Arrest	A24	No Apparent Device Problem		
							F02	Death	A24	No Apparent Device Problem		
13116	2023	All other reportable incidents	17882	Physiologic-monitoring defibrillation system	CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A0908	Incorrect, Inadequate or Imprecise Result or Readings		
							F27	Problem Identified During Non-Clinical Procedure	A0908	Incorrect, Inadequate or Imprecise Result or Readings		
13117	2023	Death	17882	Physiologic-monitoring defibrillation system	CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070908	Failure to Sense		
							F02	Death	A070908	Failure to Sense		
13118	2023	All other reportable incidents	47267	Single-chamber implantable pacemaker, rate-responsive	CT141	Electrodes/Leads and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem	19/04/2023	19/04/2023
									A070504	Premature Discharge of Battery	19/04/2023	19/04/2023
							F08	Hospitalization or Prolonged Hospitalization	A0705	Battery Problem	19/04/2023	19/04/2023
									A070504	Premature Discharge of Battery	19/04/2023	19/04/2023
							F1905	Device Revision or Replacement	A0705	Battery Problem	19/04/2023	19/04/2023
									A070504	Premature Discharge of Battery	19/04/2023	19/04/2023
					CT145	Cardiovascular devices	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem	19/04/2023	19/04/2023
									A070504	Premature Discharge of Battery	19/04/2023	19/04/2023
							F08	Hospitalization or Prolonged Hospitalization	A0705	Battery Problem	19/04/2023	19/04/2023
									A070504	Premature Discharge of Battery	19/04/2023	19/04/2023
							F1905	Device Revision or Replacement	A0705	Battery Problem	19/04/2023	19/04/2023
									A070504	Premature Discharge of Battery	19/04/2023	19/04/2023
					CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem	19/04/2023	19/04/2023
									A070504	Premature Discharge of Battery	19/04/2023	19/04/2023
							F08	Hospitalization or Prolonged Hospitalization	A0705	Battery Problem	19/04/2023	19/04/2023
									A070504	Premature Discharge of Battery	19/04/2023	19/04/2023
							F1905	Device Revision or Replacement	A0705	Battery Problem	19/04/2023	19/04/2023
									A070504	Premature Discharge of Battery	19/04/2023	19/04/2023
					CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem	19/04/2023	19/04/2023
									A070504	Premature Discharge of Battery	19/04/2023	19/04/2023
							F08	Hospitalization or Prolonged Hospitalization	A0705	Battery Problem	19/04/2023	19/04/2023
									A070504	Premature Discharge of Battery	19/04/2023	19/04/2023
							F1905	Device Revision or Replacement	A0705	Battery Problem	19/04/2023	19/04/2023
									A070504	Premature Discharge of Battery	19/04/2023	19/04/2023
					CT233	Surgical	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem	19/04/2023	19/04/2023
									A070504	Premature Discharge of Battery	19/04/2023	19/04/2023
							F08	Hospitalization or Prolonged Hospitalization	A0705	Battery Problem	19/04/2023	19/04/2023
									A070504	Premature Discharge of Battery	19/04/2023	19/04/2023
							F1905	Device Revision or Replacement	A0705	Battery Problem	19/04/2023	19/04/2023
									A070504	Premature Discharge of Battery	19/04/2023	19/04/2023
					CT2406	Implantable	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem	19/04/2023	19/04/2023
									A070504	Premature Discharge of Battery	19/04/2023	19/04/2023
							F08	Hospitalization or Prolonged Hospitalization	A0705	Battery Problem	19/04/2023	19/04/2023
									A070504	Premature Discharge of Battery	19/04/2023	19/04/2023
							F1905	Device Revision or Replacement	A0705	Battery Problem	19/04/2023	19/04/2023
									A070504	Premature Discharge of Battery	19/04/2023	19/04/2023
					CT2407	Active implantable	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem	19/04/2023	19/04/2023
									A070504	Premature Discharge of Battery	19/04/2023	19/04/2023
							F08	Hospitalization or Prolonged Hospitalization	A0705	Battery Problem	19/04/2023	19/04/2023
									A070504	Premature Discharge of Battery	19/04/2023	19/04/2023
							F1905	Device Revision or Replacement	A0705	Battery Problem	19/04/2023	19/04/2023
									A070504	Premature Discharge of Battery	19/04/2023	19/04/2023
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem	19/04/2023	19/04/2023
									A070504	Premature Discharge of Battery	19/04/2023	19/04/2023
							F08	Hospitalization or Prolonged Hospitalization	A0705	Battery Problem	19/04/2023	19/04/2023
									A070504	Premature Discharge of Battery	19/04/2023	19/04/2023
							F1905	Device Revision or Replacement	A0705	Battery Problem	19/04/2023	19/04/2023
									A070504	Premature Discharge of Battery	19/04/2023	19/04/2023
					CT2727	Active implant devices/ancillaries	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem	19/04/2023	19/04/2023
									A070504	Premature Discharge of Battery	19/04/2023	19/04/2023
							F08	Hospitalization or Prolonged Hospitalization	A0705	Battery Problem	19/04/2023	19/04/2023
									A070504	Premature Discharge of Battery	19/04/2023	19/04/2023
							F1905	Device Revision or Replacement	A0705	Battery Problem	19/04/2023	19/04/2023
									A070504	Premature Discharge of Battery	19/04/2023	19/04/2023
					CT2750	Implantable cardiac electrical stimulators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem	19/04/2023	19/04/2023
									A070504	Premature Discharge of Battery	19/04/2023	19/04/2023
							F08	Hospitalization or Prolonged Hospitalization	A0705	Battery Problem	19/04/2023	19/04/2023
									A070504	Premature Discharge of Battery	19/04/2023	19/04/2023
							F1905	Device Revision or Replacement	A0705	Battery Problem	19/04/2023	19/04/2023
									A070504	Premature Discharge of Battery	19/04/2023	19/04/2023
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem	19/04/2023	19/04/2023
									A070504	Premature Discharge of Battery	19/04/2023	19/04/2023
							F08	Hospitalization or Prolonged Hospitalization	A0705	Battery Problem	19/04/2023	19/04/2023
									A070504	Premature Discharge of Battery	19/04/2023	19/04/2023
							F1905	Device Revision or Replacement	A0705	Battery Problem	19/04/2023	19/04/2023
									A070504	Premature Discharge of Battery	19/04/2023	19/04/2023
					CT2985	Active	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem	19/04/2023	19/04/2023
									A070504	Premature Discharge of Battery	19/04/2023	19/04/2023

							F08	Hospitalization or Prolonged Hospitalization	A0705	Battery Problem	19/04/2023	19/04/2023
									A070504	Premature Discharge of Battery	19/04/2023	19/04/2023
							F1905	Device Revision or Replacement	A0705	Battery Problem	19/04/2023	19/04/2023
									A070504	Premature Discharge of Battery	19/04/2023	19/04/2023
					CT3001	Implantable metal	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem	19/04/2023	19/04/2023
									A070504	Premature Discharge of Battery	19/04/2023	19/04/2023
							F08	Hospitalization or Prolonged Hospitalization	A0705	Battery Problem	19/04/2023	19/04/2023
									A070504	Premature Discharge of Battery	19/04/2023	19/04/2023
							F1905	Device Revision or Replacement	A0705	Battery Problem	19/04/2023	19/04/2023
									A070504	Premature Discharge of Battery	19/04/2023	19/04/2023
					CT3002	Implantable synthetic polymer	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem	19/04/2023	19/04/2023
									A070504	Premature Discharge of Battery	19/04/2023	19/04/2023
							F08	Hospitalization or Prolonged Hospitalization	A0705	Battery Problem	19/04/2023	19/04/2023
									A070504	Premature Discharge of Battery	19/04/2023	19/04/2023
							F1905	Device Revision or Replacement	A0705	Battery Problem	19/04/2023	19/04/2023
									A070504	Premature Discharge of Battery	19/04/2023	19/04/2023
					CT3006	Implantable inorganic material	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem	19/04/2023	19/04/2023
									A070504	Premature Discharge of Battery	19/04/2023	19/04/2023
							F08	Hospitalization or Prolonged Hospitalization	A0705	Battery Problem	19/04/2023	19/04/2023
									A070504	Premature Discharge of Battery	19/04/2023	19/04/2023
							F1905	Device Revision or Replacement	A0705	Battery Problem	19/04/2023	19/04/2023
									A070504	Premature Discharge of Battery	19/04/2023	19/04/2023
					CT314	Electrically powered	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem	19/04/2023	19/04/2023
									A070504	Premature Discharge of Battery	19/04/2023	19/04/2023
							F08	Hospitalization or Prolonged Hospitalization	A0705	Battery Problem	19/04/2023	19/04/2023
									A070504	Premature Discharge of Battery	19/04/2023	19/04/2023
							F1905	Device Revision or Replacement	A0705	Battery Problem	19/04/2023	19/04/2023
									A070504	Premature Discharge of Battery	19/04/2023	19/04/2023
					CT321	Long-term surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem	19/04/2023	19/04/2023
									A070504	Premature Discharge of Battery	19/04/2023	19/04/2023
							F08	Hospitalization or Prolonged Hospitalization	A0705	Battery Problem	19/04/2023	19/04/2023
									A070504	Premature Discharge of Battery	19/04/2023	19/04/2023
							F1905	Device Revision or Replacement	A0705	Battery Problem	19/04/2023	19/04/2023
									A070504	Premature Discharge of Battery	19/04/2023	19/04/2023
					CT449	Generators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem	19/04/2023	19/04/2023
									A070504	Premature Discharge of Battery	19/04/2023	19/04/2023
							F08	Hospitalization or Prolonged Hospitalization	A0705	Battery Problem	19/04/2023	19/04/2023
									A070504	Premature Discharge of Battery	19/04/2023	19/04/2023
							F1905	Device Revision or Replacement	A0705	Battery Problem	19/04/2023	19/04/2023
									A070504	Premature Discharge of Battery	19/04/2023	19/04/2023
					CT981	Single use	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem	19/04/2023	19/04/2023
									A070504	Premature Discharge of Battery	19/04/2023	19/04/2023
							F08	Hospitalization or Prolonged Hospitalization	A0705	Battery Problem	19/04/2023	19/04/2023
									A070504	Premature Discharge of Battery	19/04/2023	19/04/2023
							F1905	Device Revision or Replacement	A0705	Battery Problem	19/04/2023	19/04/2023
									A070504	Premature Discharge of Battery	19/04/2023	19/04/2023
					CT983	Surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem	19/04/2023	19/04/2023
									A070504	Premature Discharge of Battery	19/04/2023	19/04/2023
							F08	Hospitalization or Prolonged Hospitalization	A0705	Battery Problem	19/04/2023	19/04/2023
									A070504	Premature Discharge of Battery	19/04/2023	19/04/2023
							F1905	Device Revision or Replacement	A0705	Battery Problem	19/04/2023	19/04/2023
									A070504	Premature Discharge of Battery	19/04/2023	19/04/2023
13119	2023	All other reportable incidents	47265	Dual-chamber implantable pacemaker, rate-responsive	CT1685	Cardiac pacemakers	E011903	Syncope/Fainting	A070909	Oversensing	15/01/2023	15/01/2023
									A0712	Pacing Problem	15/01/2023	15/01/2023
									A0908	Incorrect, Inadequate or Imprecise Result or Readings	15/01/2023	15/01/2023
									A090801	Signal Artifact/Noise	15/01/2023	15/01/2023
							F08	Hospitalization or Prolonged Hospitalization	A070909	Oversensing	15/01/2023	15/01/2023
									A0712	Pacing Problem	15/01/2023	15/01/2023
									A0908	Incorrect, Inadequate or Imprecise Result or Readings	15/01/2023	15/01/2023
									A090801	Signal Artifact/Noise	15/01/2023	15/01/2023
							F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A070909	Oversensing	15/01/2023	15/01/2023
									A0712	Pacing Problem	15/01/2023	15/01/2023
									A0908	Incorrect, Inadequate or Imprecise Result or Readings	15/01/2023	15/01/2023
									A090801	Signal Artifact/Noise	15/01/2023	15/01/2023
							F1905	Device Revision or Replacement	A070909	Oversensing	15/01/2023	15/01/2023
									A0712	Pacing Problem	15/01/2023	15/01/2023
									A0908	Incorrect, Inadequate or Imprecise Result or Readings	15/01/2023	15/01/2023
									A090801	Signal Artifact/Noise	15/01/2023	15/01/2023
					CT257	Implantable cardiac pacemakers	E011903	Syncope/Fainting	A070909	Oversensing	15/01/2023	15/01/2023
									A0712	Pacing Problem	15/01/2023	15/01/2023
									A0908	Incorrect, Inadequate or Imprecise Result or Readings	15/01/2023	15/01/2023
									A090801	Signal Artifact/Noise	15/01/2023	15/01/2023
							F08	Hospitalization or Prolonged Hospitalization	A070909	Oversensing	15/01/2023	15/01/2023
									A0712	Pacing Problem	15/01/2023	15/01/2023
									A0908	Incorrect, Inadequate or Imprecise Result or Readings	15/01/2023	15/01/2023
									A090801	Signal Artifact/Noise	15/01/2023	15/01/2023
							F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A070909	Oversensing	15/01/2023	15/01/2023
									A0712	Pacing Problem	15/01/2023	15/01/2023
									A0908	Incorrect, Inadequate or Imprecise Result or Readings	15/01/2023	15/01/2023
									A090801	Signal Artifact/Noise	15/01/2023	15/01/2023
							F1905	Device Revision or Replacement	A070909	Oversensing	15/01/2023	15/01/2023
									A0712	Pacing Problem	15/01/2023	15/01/2023

								A0908	Incorrect, Inadequate or Imprecise Result or Readings	15/01/2023	15/01/2023
								A090801	Signal Artifact/Noise	15/01/2023	15/01/2023
					CT2866	Implantable cardiac electrical stimulators	E011903	Syncope/Fainting	A070909	Oversensing	15/01/2023	15/01/2023
									A0712	Pacing Problem	15/01/2023	15/01/2023
									A0908	Incorrect, Inadequate or Imprecise Result or Readings	15/01/2023	15/01/2023
									A090801	Signal Artifact/Noise	15/01/2023	15/01/2023
							F08	Hospitalization or Prolonged Hospitalization	A070909	Oversensing	15/01/2023	15/01/2023
									A0712	Pacing Problem	15/01/2023	15/01/2023
									A0908	Incorrect, Inadequate or Imprecise Result or Readings	15/01/2023	15/01/2023
									A090801	Signal Artifact/Noise	15/01/2023	15/01/2023
							F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A070909	Oversensing	15/01/2023	15/01/2023
									A0712	Pacing Problem	15/01/2023	15/01/2023
									A0908	Incorrect, Inadequate or Imprecise Result or Readings	15/01/2023	15/01/2023
									A090801	Signal Artifact/Noise	15/01/2023	15/01/2023
							F1905	Device Revision or Replacement	A070909	Oversensing	15/01/2023	15/01/2023
									A0712	Pacing Problem	15/01/2023	15/01/2023
									A0908	Incorrect, Inadequate or Imprecise Result or Readings	15/01/2023	15/01/2023
									A090801	Signal Artifact/Noise	15/01/2023	15/01/2023
13120	2023	All other reportable incidents	35852	Single-chamber implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	20/06/2023	20/06/2023
									A0908	Incorrect, Inadequate or Imprecise Result or Readings	20/06/2023	20/06/2023
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	20/06/2023	20/06/2023
									A0908	Incorrect, Inadequate or Imprecise Result or Readings	20/06/2023	20/06/2023
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	20/06/2023	20/06/2023
									A0908	Incorrect, Inadequate or Imprecise Result or Readings	20/06/2023	20/06/2023
13121	2023	All other reportable incidents	36241	Pacemaker/defibrillator lead	CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070910	Undersensing		
									A0712	Pacing Problem		
									A0909	Unexpected Therapeutic Results		
							F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A070910	Undersensing		
									A0712	Pacing Problem		
									A0909	Unexpected Therapeutic Results		
13122	2023	All other reportable incidents	47265	Dual-chamber implantable pacemaker, rate-responsive	CT141	Electrodes/Leads and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
					CT145	Cardiovascular devices	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
					CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
					CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
					CT233	Surgical	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
					CT2406	Implantable	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
					CT2407	Active implantable	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
					CT2727	Active implant devices/ancillaries	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
					CT2750	Implantable cardiac electrical stimulators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
					CT2985	Active	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
					CT3001	Implantable metal	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		

					CT3002	Implantable synthetic polymer	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
					CT3006	Implantable inorganic material	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
					CT314	Electrically powered	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
					CT321	Long-term surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
					CT449	Generators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
					CT981	Single use	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
					CT983	Surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
13123	2023	All other reportable incidents	47263	Cardiac resynchronization therapy implantable pacemaker	CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A090805	Incorrect Measurement		
							F23	Unexpected Medical Intervention	A090805	Incorrect Measurement		
					CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A090805	Incorrect Measurement		
							F23	Unexpected Medical Intervention	A090805	Incorrect Measurement		
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A090805	Incorrect Measurement		
							F23	Unexpected Medical Intervention	A090805	Incorrect Measurement		
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A090805	Incorrect Measurement		
							F23	Unexpected Medical Intervention	A090805	Incorrect Measurement		
13124	2023	All other reportable incidents	17882	Physiologic-monitoring defibrillation system	CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070803	Failure to Power Up		
							F27	Problem Identified During Non-Clinical Procedure	A070803	Failure to Power Up		
13125	2023	All other reportable incidents	35223	Endocardial/interventricular septal pacing lead	CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070101	Failure to Capture		
									A072201	High Impedance		
							F26	No Health Consequences or Impact	A070101	Failure to Capture		
									A072201	High Impedance		
13126	2023	All other reportable incidents	35223	Endocardial/interventricular septal pacing lead	CT1684	Cardiac pacemakers and associated devices	E0601	Arrhythmia	A07	Electrical /Electronic Property Problem		
									A070909	Oversensing		
							E060109	Tachycardia	A07	Electrical /Electronic Property Problem		
									A070909	Oversensing		
							F11	Non-Serious Injury/ Illness/ Impairment	A07	Electrical /Electronic Property Problem		
									A070909	Oversensing		
13127	2004		35852	Single-chamber implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	F1903	Device Explantation	A0406	Material Deformation		
13128	2023	All other reportable incidents	35224	External pacemaker, epicardial pacing	CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem		
							F26	No Health Consequences or Impact	A0705	Battery Problem		
					CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem		
							F26	No Health Consequences or Impact	A0705	Battery Problem		
13129	2023	All other reportable incidents	35852	Single-chamber implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
									A0908	Incorrect, Inadequate or Imprecise Result or Readings		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
									A0908	Incorrect, Inadequate or Imprecise Result or Readings		
13130	2023	All other reportable incidents	17882	Physiologic-monitoring defibrillation system	CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A071204	No Pacing		
							F05	Delay to Treatment/ Therapy	A071204	No Pacing		
13131	2023	All other reportable incidents	47265	Dual-chamber implantable pacemaker, rate-responsive	CT141	Electrodes/Leads and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A072202	Low Impedance		
							F26	No Health Consequences or Impact	A072202	Low Impedance		
					CT145	Cardiovascular devices	E2403	No Clinical Signs, Symptoms or Conditions	A072202	Low Impedance		
							F26	No Health Consequences or Impact	A072202	Low Impedance		
					CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A072202	Low Impedance		
							F26	No Health Consequences or Impact	A072202	Low Impedance		
					CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A072202	Low Impedance		
							F26	No Health Consequences or Impact	A072202	Low Impedance		
					CT233	Surgical	E2403	No Clinical Signs, Symptoms or Conditions	A072202	Low Impedance		
							F26	No Health Consequences or Impact	A072202	Low Impedance		
					CT2406	Implantable	E2403	No Clinical Signs, Symptoms or Conditions	A072202	Low Impedance		
							F26	No Health Consequences or Impact	A072202	Low Impedance		
					CT2407	Active implantable	E2403	No Clinical Signs, Symptoms or Conditions	A072202	Low Impedance		
							F26	No Health Consequences or Impact	A072202	Low Impedance		
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A072202	Low Impedance		
							F26	No Health Consequences or Impact	A072202	Low Impedance		
					CT2727	Active implant devices/ancillaries	E2403	No Clinical Signs, Symptoms or Conditions	A072202	Low Impedance		
							F26	No Health Consequences or Impact	A072202	Low Impedance		
					CT2750	Implantable cardiac electrical stimulators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A072202	Low Impedance		
							F26	No Health Consequences or Impact	A072202	Low Impedance		
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A072202	Low Impedance		
							F26	No Health Consequences or Impact	A072202	Low Impedance		
					CT2985	Active	E2403	No Clinical Signs, Symptoms or Conditions	A072202	Low Impedance		
							F26	No Health Consequences or Impact	A072202	Low Impedance		
					CT3001	Implantable metal	E2403	No Clinical Signs, Symptoms or Conditions	A072202	Low Impedance		
							F26	No Health Consequences or Impact	A072202	Low Impedance		
					CT3002	Implantable synthetic polymer	E2403	No Clinical Signs, Symptoms or Conditions	A072202	Low Impedance		

						F26	No Health Consequences or Impact	A072202	Low Impedance			
				CT3006	Implantable inorganic material	E2403	No Clinical Signs, Symptoms or Conditions	A072202	Low Impedance			
						F26	No Health Consequences or Impact	A072202	Low Impedance			
				CT314	Electrically powered	E2403	No Clinical Signs, Symptoms or Conditions	A072202	Low Impedance			
						F26	No Health Consequences or Impact	A072202	Low Impedance			
				CT321	Long-term surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A072202	Low Impedance			
						F26	No Health Consequences or Impact	A072202	Low Impedance			
				CT449	Generators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A072202	Low Impedance			
						F26	No Health Consequences or Impact	A072202	Low Impedance			
				CT981	Single use	E2403	No Clinical Signs, Symptoms or Conditions	A072202	Low Impedance			
						F26	No Health Consequences or Impact	A072202	Low Impedance			
				CT983	Surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A072202	Low Impedance			
						F26	No Health Consequences or Impact	A072202	Low Impedance			
13132	2023		47263	Cardiac resynchronization therapy implantable pacemaker	CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A0203	Defective Device	01/12/2022	01/12/2022
									A070909	Oversensing	01/12/2022	01/12/2022
									A0712	Pacing Problem	01/12/2022	01/12/2022
									A071204	No Pacing	01/12/2022	01/12/2022
							F1901	Additional Surgery	A0203	Defective Device	01/12/2022	01/12/2022
									A070909	Oversensing	01/12/2022	01/12/2022
									A0712	Pacing Problem	01/12/2022	01/12/2022
									A071204	No Pacing	01/12/2022	01/12/2022
					CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A0203	Defective Device	01/12/2022	01/12/2022
									A070909	Oversensing	01/12/2022	01/12/2022
									A0712	Pacing Problem	01/12/2022	01/12/2022
									A071204	No Pacing	01/12/2022	01/12/2022
							F1901	Additional Surgery	A0203	Defective Device	01/12/2022	01/12/2022
									A070909	Oversensing	01/12/2022	01/12/2022
									A0712	Pacing Problem	01/12/2022	01/12/2022
									A071204	No Pacing	01/12/2022	01/12/2022
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A0203	Defective Device	01/12/2022	01/12/2022
									A070909	Oversensing	01/12/2022	01/12/2022
									A0712	Pacing Problem	01/12/2022	01/12/2022
									A071204	No Pacing	01/12/2022	01/12/2022
							F1901	Additional Surgery	A0203	Defective Device	01/12/2022	01/12/2022
									A070909	Oversensing	01/12/2022	01/12/2022
									A0712	Pacing Problem	01/12/2022	01/12/2022
									A071204	No Pacing	01/12/2022	01/12/2022
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A0203	Defective Device	01/12/2022	01/12/2022
									A070909	Oversensing	01/12/2022	01/12/2022
									A0712	Pacing Problem	01/12/2022	01/12/2022
									A071204	No Pacing	01/12/2022	01/12/2022
							F1901	Additional Surgery	A0203	Defective Device	01/12/2022	01/12/2022
									A070909	Oversensing	01/12/2022	01/12/2022
									A0712	Pacing Problem	01/12/2022	01/12/2022
									A071204	No Pacing	01/12/2022	01/12/2022
13133	2023	All other reportable incidents	47265	Dual-chamber implantable pacemaker, rate-responsive	CT141	Electrodes/Leads and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A0203	Defective Device	01/10/2022	01/10/2022
									A072201	High Impedance	01/10/2022	01/10/2022
							F1901	Additional Surgery	A0203	Defective Device	01/10/2022	01/10/2022
									A072201	High Impedance	01/10/2022	01/10/2022
					CT145	Cardiovascular devices	E2403	No Clinical Signs, Symptoms or Conditions	A0203	Defective Device	01/10/2022	01/10/2022
									A072201	High Impedance	01/10/2022	01/10/2022
							F1901	Additional Surgery	A0203	Defective Device	01/10/2022	01/10/2022
									A072201	High Impedance	01/10/2022	01/10/2022
					CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A0203	Defective Device	01/10/2022	01/10/2022
									A072201	High Impedance	01/10/2022	01/10/2022
							F1901	Additional Surgery	A0203	Defective Device	01/10/2022	01/10/2022
									A072201	High Impedance	01/10/2022	01/10/2022
					CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A0203	Defective Device	01/10/2022	01/10/2022
									A072201	High Impedance	01/10/2022	01/10/2022
							F1901	Additional Surgery	A0203	Defective Device	01/10/2022	01/10/2022
									A072201	High Impedance	01/10/2022	01/10/2022
					CT233	Surgical	E2403	No Clinical Signs, Symptoms or Conditions	A0203	Defective Device	01/10/2022	01/10/2022
									A072201	High Impedance	01/10/2022	01/10/2022
							F1901	Additional Surgery	A0203	Defective Device	01/10/2022	01/10/2022
									A072201	High Impedance	01/10/2022	01/10/2022
					CT2406	Implantable	E2403	No Clinical Signs, Symptoms or Conditions	A0203	Defective Device	01/10/2022	01/10/2022
									A072201	High Impedance	01/10/2022	01/10/2022
							F1901	Additional Surgery	A0203	Defective Device	01/10/2022	01/10/2022
									A072201	High Impedance	01/10/2022	01/10/2022
					CT2407	Active implantable	E2403	No Clinical Signs, Symptoms or Conditions	A0203	Defective Device	01/10/2022	01/10/2022
									A072201	High Impedance	01/10/2022	01/10/2022
							F1901	Additional Surgery	A0203	Defective Device	01/10/2022	01/10/2022
									A072201	High Impedance	01/10/2022	01/10/2022
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A0203	Defective Device	01/10/2022	01/10/2022
									A072201	High Impedance	01/10/2022	01/10/2022
							F1901	Additional Surgery	A0203	Defective Device	01/10/2022	01/10/2022
									A072201	High Impedance	01/10/2022	01/10/2022
					CT2727	Active implant devices/ancillaries	E2403	No Clinical Signs, Symptoms or Conditions	A0203	Defective Device	01/10/2022	01/10/2022
									A072201	High Impedance	01/10/2022	01/10/2022
							F1901	Additional Surgery	A0203	Defective Device	01/10/2022	01/10/2022
									A072201	High Impedance	01/10/2022	01/10/2022
					CT2750	Implantable cardiac electrical stimulators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A0203	Defective Device	01/10/2022	01/10/2022
									A072201	High Impedance	01/10/2022	01/10/2022
							F1901	Additional Surgery	A0203	Defective Device	01/10/2022	01/10/2022
									A072201	High Impedance	01/10/2022	01/10/2022
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A0203	Defective Device	01/10/2022	01/10/2022
									A072201	High Impedance	01/10/2022	01/10/2022
							F1901	Additional Surgery	A0203	Defective Device	01/10/2022	01/10/2022

									A072201	High Impedance	01/10/2022	01/10/2022
					CT2985	Active	E2403	No Clinical Signs, Symptoms or Conditions	A0203	Defective Device	01/10/2022	01/10/2022
									A072201	High Impedance	01/10/2022	01/10/2022
							F1901	Additional Surgery	A0203	Defective Device	01/10/2022	01/10/2022
									A072201	High Impedance	01/10/2022	01/10/2022
					CT3001	Implantable metal	E2403	No Clinical Signs, Symptoms or Conditions	A0203	Defective Device	01/10/2022	01/10/2022
									A072201	High Impedance	01/10/2022	01/10/2022
							F1901	Additional Surgery	A0203	Defective Device	01/10/2022	01/10/2022
									A072201	High Impedance	01/10/2022	01/10/2022
					CT3002	Implantable synthetic polymer	E2403	No Clinical Signs, Symptoms or Conditions	A0203	Defective Device	01/10/2022	01/10/2022
									A072201	High Impedance	01/10/2022	01/10/2022
							F1901	Additional Surgery	A0203	Defective Device	01/10/2022	01/10/2022
									A072201	High Impedance	01/10/2022	01/10/2022
					CT3006	Implantable inorganic material	E2403	No Clinical Signs, Symptoms or Conditions	A0203	Defective Device	01/10/2022	01/10/2022
									A072201	High Impedance	01/10/2022	01/10/2022
							F1901	Additional Surgery	A0203	Defective Device	01/10/2022	01/10/2022
									A072201	High Impedance	01/10/2022	01/10/2022
					CT314	Electrically powered	E2403	No Clinical Signs, Symptoms or Conditions	A0203	Defective Device	01/10/2022	01/10/2022
									A072201	High Impedance	01/10/2022	01/10/2022
							F1901	Additional Surgery	A0203	Defective Device	01/10/2022	01/10/2022
									A072201	High Impedance	01/10/2022	01/10/2022
					CT321	Long-term surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A0203	Defective Device	01/10/2022	01/10/2022
									A072201	High Impedance	01/10/2022	01/10/2022
							F1901	Additional Surgery	A0203	Defective Device	01/10/2022	01/10/2022
									A072201	High Impedance	01/10/2022	01/10/2022
					CT449	Generators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A0203	Defective Device	01/10/2022	01/10/2022
									A072201	High Impedance	01/10/2022	01/10/2022
							F1901	Additional Surgery	A0203	Defective Device	01/10/2022	01/10/2022
									A072201	High Impedance	01/10/2022	01/10/2022
					CT981	Single use	E2403	No Clinical Signs, Symptoms or Conditions	A0203	Defective Device	01/10/2022	01/10/2022
									A072201	High Impedance	01/10/2022	01/10/2022
							F1901	Additional Surgery	A0203	Defective Device	01/10/2022	01/10/2022
									A072201	High Impedance	01/10/2022	01/10/2022
					CT983	Surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A0203	Defective Device	01/10/2022	01/10/2022
									A072201	High Impedance	01/10/2022	01/10/2022
							F1901	Additional Surgery	A0203	Defective Device	01/10/2022	01/10/2022
									A072201	High Impedance	01/10/2022	01/10/2022
13134	2023	All other reportable incidents	17882	Physiologic-monitoring defibrillation system	CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A090402	Failure to Deliver Energy		
							F26	No Health Consequences or Impact	A090402	Failure to Deliver Energy		
13135	2023	All other reportable incidents	47205	Cardiac pulse generator programmer	CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A090805	Incorrect Measurement		
							F23	Unexpected Medical Intervention	A090805	Incorrect Measurement		
13136	2023	All other reportable incidents	47270	Cardiac resynchronization therapy implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A070910	Undersensing		
							F23	Unexpected Medical Intervention	A070910	Undersensing		
13137	2023	All other reportable incidents	36241	Pacemaker/defibrillator lead	CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A072201	High Impedance		
							F26	No Health Consequences or Impact	A072201	High Impedance		
13138	2023	All other reportable incidents	37265	Dual-chamber implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A072201	High Impedance		
							F26	No Health Consequences or Impact	A072201	High Impedance		
13139	2023	All other reportable incidents	47265	Dual-chamber implantable pacemaker, rate-responsive	CT141	Electrodes/Leads and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
							F24	Insufficient Health Impact Information	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
					CT145	Cardiovascular devices	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
							F24	Insufficient Health Impact Information	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
					CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
							F24	Insufficient Health Impact Information	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
					CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
							F24	Insufficient Health Impact Information	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
					CT233	Surgical	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
							F24	Insufficient Health Impact Information	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
					CT2406	Implantable	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
							F24	Insufficient Health Impact Information	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
					CT2407	Active implantable	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
							F24	Insufficient Health Impact Information	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
							F24	Insufficient Health Impact Information	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
					CT2727	Active implant devices/ancillaries	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
							F24	Insufficient Health Impact Information	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
					CT2750	Implantable cardiac electrical stimulators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
							F24	Insufficient Health Impact Information	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem		

									A070504	Premature Discharge of Battery		
							F24	Insufficient Health Impact Information	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
					CT2985	Active	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
							F24	Insufficient Health Impact Information	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
					CT3001	Implantable metal	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
							F24	Insufficient Health Impact Information	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
					CT3002	Implantable synthetic polymer	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
							F24	Insufficient Health Impact Information	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
					CT3006	Implantable inorganic material	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
							F24	Insufficient Health Impact Information	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
					CT314	Electrically powered	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
							F24	Insufficient Health Impact Information	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
					CT321	Long-term surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
							F24	Insufficient Health Impact Information	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
					CT449	Generators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
							F24	Insufficient Health Impact Information	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
					CT981	Single use	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
							F24	Insufficient Health Impact Information	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
					CT983	Surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
							F24	Insufficient Health Impact Information	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
13140	2023	All other reportable incidents	47265	Dual-chamber implantable pacemaker, rate-responsive	CT141	Electrodes/Leads and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	12/05/2023	12/05/2023
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	12/05/2023	12/05/2023
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	12/05/2023	12/05/2023
					CT145	Cardiovascular devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	12/05/2023	12/05/2023
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	12/05/2023	12/05/2023
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	12/05/2023	12/05/2023
					CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	12/05/2023	12/05/2023
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	12/05/2023	12/05/2023
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	12/05/2023	12/05/2023
					CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	12/05/2023	12/05/2023
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	12/05/2023	12/05/2023
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	12/05/2023	12/05/2023
					CT233	Surgical	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	12/05/2023	12/05/2023
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	12/05/2023	12/05/2023
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	12/05/2023	12/05/2023
					CT2406	Implantable	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	12/05/2023	12/05/2023
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	12/05/2023	12/05/2023
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	12/05/2023	12/05/2023
					CT2407	Active implantable	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	12/05/2023	12/05/2023
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	12/05/2023	12/05/2023
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	12/05/2023	12/05/2023
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	12/05/2023	12/05/2023
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	12/05/2023	12/05/2023
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	12/05/2023	12/05/2023
					CT2727	Active implant devices/ancillaries	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	12/05/2023	12/05/2023
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	12/05/2023	12/05/2023
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	12/05/2023	12/05/2023
					CT2750	Implantable cardiac electrical stimulators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	12/05/2023	12/05/2023
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	12/05/2023	12/05/2023
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	12/05/2023	12/05/2023
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	12/05/2023	12/05/2023
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	12/05/2023	12/05/2023
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	12/05/2023	12/05/2023
					CT2985	Active	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	12/05/2023	12/05/2023
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	12/05/2023	12/05/2023
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	12/05/2023	12/05/2023
					CT3001	Implantable metal	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	12/05/2023	12/05/2023
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	12/05/2023	12/05/2023
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	12/05/2023	12/05/2023
					CT3002	Implantable synthetic polymer	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	12/05/2023	12/05/2023
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	12/05/2023	12/05/2023
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	12/05/2023	12/05/2023
					CT3006	Implantable inorganic material	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	12/05/2023	12/05/2023
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	12/05/2023	12/05/2023
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	12/05/2023	12/05/2023
					CT314	Electrically powered	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	12/05/2023	12/05/2023
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	12/05/2023	12/05/2023
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	12/05/2023	12/05/2023
					CT321	Long-term surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	12/05/2023	12/05/2023

						F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	12/05/2023	12/05/2023	
						F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	12/05/2023	12/05/2023	
					CT449	Generators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	12/05/2023	12/05/2023
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	12/05/2023	12/05/2023
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	12/05/2023	12/05/2023
					CT981	Single use	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	12/05/2023	12/05/2023
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	12/05/2023	12/05/2023
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	12/05/2023	12/05/2023
					CT983	Surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	12/05/2023	12/05/2023
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	12/05/2023	12/05/2023
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	12/05/2023	12/05/2023
13141	2023	All other reportable incidents	35852	Single-chamber implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem	12/04/2023	12/04/2023
									A070504	Premature Discharge of Battery	12/04/2023	12/04/2023
							F08	Hospitalization or Prolonged Hospitalization	A0705	Battery Problem	12/04/2023	12/04/2023
									A070504	Premature Discharge of Battery	12/04/2023	12/04/2023
							F1903	Device Explantation	A0705	Battery Problem	12/04/2023	12/04/2023
									A070504	Premature Discharge of Battery	12/04/2023	12/04/2023
13142	2023		47265	Dual-chamber implantable pacemaker, rate-responsive	CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	27/02/2023	27/02/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	27/02/2023	27/02/2023
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	27/02/2023	27/02/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	27/02/2023	27/02/2023
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	27/02/2023	27/02/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	27/02/2023	27/02/2023
13143	2023	All other reportable incidents	47267	Single-chamber implantable pacemaker, rate-responsive	CT141	Electrodes/Leads and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT145	Cardiovascular devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT233	Surgical	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT2406	Implantable	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT2407	Active implantable	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT2727	Active implant devices/ancillaries	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT2750	Implantable cardiac electrical stimulators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT2985	Active	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT3001	Implantable metal	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT3002	Implantable synthetic polymer	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT3006	Implantable inorganic material	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT314	Electrically powered	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT321	Long-term surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT449	Generators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT981	Single use	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT983	Surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
13144	2023	All other reportable incidents	47265	Dual-chamber implantable pacemaker, rate-responsive	CT141	Electrodes/Leads and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
					CT145	Cardiovascular devices	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
					CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
					CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
					CT233	Surgical	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
					CT2406	Implantable	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
					CT2407	Active implantable	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem		

									A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
					CT2727	Active implant devices/ancillaries	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
					CT2750	Implantable cardiac electrical stimulators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
					CT2985	Active	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
					CT3001	Implantable metal	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
					CT3002	Implantable synthetic polymer	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
					CT3006	Implantable inorganic material	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
					CT314	Electrically powered	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
					CT321	Long-term surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
					CT449	Generators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
					CT981	Single use	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
					CT983	Surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
13145	2023	All other reportable incidents	37265	Dual-chamber implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A071302	Inappropriate/Inadequate Shock/Stimulation		
							F12	Serious Injury/ Illness/ Impairment	A071302	Inappropriate/Inadequate Shock/Stimulation		
13146	2023	All other reportable incidents	35852	Single-chamber implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A0713	Defibrillation/Stimulation Problem		
							F12	Serious Injury/ Illness/ Impairment	A0713	Defibrillation/Stimulation Problem		
							F23	Unexpected Medical Intervention	A0713	Defibrillation/Stimulation Problem		
13147	2023		47265	Dual-chamber implantable pacemaker, rate-responsive	CT141	Electrodes/Leads and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	03/05/2023	03/05/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	03/05/2023	03/05/2023
					CT145	Cardiovascular devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	03/05/2023	03/05/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	03/05/2023	03/05/2023
					CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	03/05/2023	03/05/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	03/05/2023	03/05/2023
					CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	03/05/2023	03/05/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	03/05/2023	03/05/2023
					CT233	Surgical	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	03/05/2023	03/05/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	03/05/2023	03/05/2023
					CT2406	Implantable	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	03/05/2023	03/05/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	03/05/2023	03/05/2023
					CT2407	Active implantable	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	03/05/2023	03/05/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	03/05/2023	03/05/2023
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	03/05/2023	03/05/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	03/05/2023	03/05/2023
					CT2727	Active implant devices/ancillaries	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	03/05/2023	03/05/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	03/05/2023	03/05/2023
					CT2750	Implantable cardiac electrical stimulators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	03/05/2023	03/05/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	03/05/2023	03/05/2023
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	03/05/2023	03/05/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	03/05/2023	03/05/2023
					CT2985	Active	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	03/05/2023	03/05/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	03/05/2023	03/05/2023
					CT3001	Implantable metal	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	03/05/2023	03/05/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	03/05/2023	03/05/2023
					CT3002	Implantable synthetic polymer	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	03/05/2023	03/05/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	03/05/2023	03/05/2023

					CT3006	Implantable inorganic material	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	03/05/2023	03/05/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	03/05/2023	03/05/2023
					CT314	Electrically powered	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	03/05/2023	03/05/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	03/05/2023	03/05/2023
					CT321	Long-term surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	03/05/2023	03/05/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	03/05/2023	03/05/2023
					CT449	Generators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	03/05/2023	03/05/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	03/05/2023	03/05/2023
					CT981	Single use	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	03/05/2023	03/05/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	03/05/2023	03/05/2023
					CT983	Surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	03/05/2023	03/05/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	03/05/2023	03/05/2023
13148	2023		47265	Dual-chamber implantable pacemaker, rate-responsive	CT141	Electrodes/Leads and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	03/05/2023	03/05/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	03/05/2023	03/05/2023
					CT145	Cardiovascular devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	03/05/2023	03/05/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	03/05/2023	03/05/2023
					CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	03/05/2023	03/05/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	03/05/2023	03/05/2023
					CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	03/05/2023	03/05/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	03/05/2023	03/05/2023
					CT233	Surgical	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	03/05/2023	03/05/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	03/05/2023	03/05/2023
					CT2406	Implantable	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	03/05/2023	03/05/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	03/05/2023	03/05/2023
					CT2407	Active implantable	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	03/05/2023	03/05/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	03/05/2023	03/05/2023
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	03/05/2023	03/05/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	03/05/2023	03/05/2023
					CT2727	Active implant devices/ancillaries	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	03/05/2023	03/05/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	03/05/2023	03/05/2023
					CT2750	Implantable cardiac electrical stimulators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	03/05/2023	03/05/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	03/05/2023	03/05/2023
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	03/05/2023	03/05/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	03/05/2023	03/05/2023
					CT2985	Active	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	03/05/2023	03/05/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	03/05/2023	03/05/2023
					CT3001	Implantable metal	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	03/05/2023	03/05/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	03/05/2023	03/05/2023
					CT3002	Implantable synthetic polymer	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	03/05/2023	03/05/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	03/05/2023	03/05/2023
					CT3006	Implantable inorganic material	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	03/05/2023	03/05/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	03/05/2023	03/05/2023
					CT314	Electrically powered	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	03/05/2023	03/05/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	03/05/2023	03/05/2023
					CT321	Long-term surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	03/05/2023	03/05/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	03/05/2023	03/05/2023
					CT449	Generators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	03/05/2023	03/05/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	03/05/2023	03/05/2023
					CT981	Single use	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	03/05/2023	03/05/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	03/05/2023	03/05/2023
					CT983	Surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	03/05/2023	03/05/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	03/05/2023	03/05/2023
13149	2023	All other reportable incidents	47267	Single-chamber implantable pacemaker, rate-responsive	CT141	Electrodes/Leads and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT145	Cardiovascular devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT233	Surgical	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT2406	Implantable	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT2407	Active implantable	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT2727	Active implant devices/ancillaries	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT2750	Implantable cardiac electrical stimulators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT2985	Active	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT3001	Implantable metal	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT3002	Implantable synthetic polymer	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT3006	Implantable inorganic material	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT314	Electrically powered	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT321	Long-term surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT449	Generators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		

					CT981	Single use	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT983	Surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
13150	2023	All other reportable incidents	17882	Physiologic-monitoring defibrillation system	CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A1603	Failure of Device to Self-Test		
							F27	Problem Identified During Non-Clinical Procedure	A1603	Failure of Device to Self-Test		
13151	2023	All other reportable incidents	36241	Pacemaker/defibrillator lead	CT1684	Cardiac pacemakers and associated devices	E0610	Endocarditis	A24	No Apparent Device Problem	10/05/2023	10/05/2023
							F08	Hospitalization or Prolonged Hospitalization	A24	No Apparent Device Problem	10/05/2023	10/05/2023
							F1903	Device Explantation	A24	No Apparent Device Problem	10/05/2023	10/05/2023
							F2203	Imaging Required	A24	No Apparent Device Problem	10/05/2023	10/05/2023
13152	2023	All other reportable incidents	17882	Physiologic-monitoring defibrillation system	CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A071301	Failure to Deliver Shock/Stimulation		
									A0719	Unexpected Shutdown		
							F26	No Health Consequences or Impact	A071301	Failure to Deliver Shock/Stimulation		
									A0719	Unexpected Shutdown		
13153	2023		47265	Dual-chamber implantable pacemaker, rate-responsive	CT141	Electrodes/Leads and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	19/06/2023	19/06/2023
									A090805	Incorrect Measurement	19/06/2023	19/06/2023
							F1901	Additional Surgery	A070504	Premature Discharge of Battery	19/06/2023	19/06/2023
									A090805	Incorrect Measurement	19/06/2023	19/06/2023
					CT145	Cardiovascular devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	19/06/2023	19/06/2023
									A090805	Incorrect Measurement	19/06/2023	19/06/2023
							F1901	Additional Surgery	A070504	Premature Discharge of Battery	19/06/2023	19/06/2023
									A090805	Incorrect Measurement	19/06/2023	19/06/2023
					CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	19/06/2023	19/06/2023
									A090805	Incorrect Measurement	19/06/2023	19/06/2023
							F1901	Additional Surgery	A070504	Premature Discharge of Battery	19/06/2023	19/06/2023
									A090805	Incorrect Measurement	19/06/2023	19/06/2023
					CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	19/06/2023	19/06/2023
									A090805	Incorrect Measurement	19/06/2023	19/06/2023
							F1901	Additional Surgery	A070504	Premature Discharge of Battery	19/06/2023	19/06/2023
									A090805	Incorrect Measurement	19/06/2023	19/06/2023
					CT233	Surgical	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	19/06/2023	19/06/2023
									A090805	Incorrect Measurement	19/06/2023	19/06/2023
							F1901	Additional Surgery	A070504	Premature Discharge of Battery	19/06/2023	19/06/2023
									A090805	Incorrect Measurement	19/06/2023	19/06/2023
					CT2406	Implantable	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	19/06/2023	19/06/2023
									A090805	Incorrect Measurement	19/06/2023	19/06/2023
							F1901	Additional Surgery	A070504	Premature Discharge of Battery	19/06/2023	19/06/2023
									A090805	Incorrect Measurement	19/06/2023	19/06/2023
					CT2407	Active implantable	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	19/06/2023	19/06/2023
									A090805	Incorrect Measurement	19/06/2023	19/06/2023
							F1901	Additional Surgery	A070504	Premature Discharge of Battery	19/06/2023	19/06/2023
									A090805	Incorrect Measurement	19/06/2023	19/06/2023
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	19/06/2023	19/06/2023
									A090805	Incorrect Measurement	19/06/2023	19/06/2023
							F1901	Additional Surgery	A070504	Premature Discharge of Battery	19/06/2023	19/06/2023
									A090805	Incorrect Measurement	19/06/2023	19/06/2023
					CT2727	Active implant devices/ancillaries	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	19/06/2023	19/06/2023
									A090805	Incorrect Measurement	19/06/2023	19/06/2023
							F1901	Additional Surgery	A070504	Premature Discharge of Battery	19/06/2023	19/06/2023
									A090805	Incorrect Measurement	19/06/2023	19/06/2023
					CT2750	Implantable cardiac electrical stimulators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	19/06/2023	19/06/2023
									A090805	Incorrect Measurement	19/06/2023	19/06/2023
							F1901	Additional Surgery	A070504	Premature Discharge of Battery	19/06/2023	19/06/2023
									A090805	Incorrect Measurement	19/06/2023	19/06/2023
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	19/06/2023	19/06/2023
									A090805	Incorrect Measurement	19/06/2023	19/06/2023
							F1901	Additional Surgery	A070504	Premature Discharge of Battery	19/06/2023	19/06/2023
									A090805	Incorrect Measurement	19/06/2023	19/06/2023
					CT2985	Active	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	19/06/2023	19/06/2023
									A090805	Incorrect Measurement	19/06/2023	19/06/2023
							F1901	Additional Surgery	A070504	Premature Discharge of Battery	19/06/2023	19/06/2023
									A090805	Incorrect Measurement	19/06/2023	19/06/2023
					CT3001	Implantable metal	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	19/06/2023	19/06/2023
									A090805	Incorrect Measurement	19/06/2023	19/06/2023
							F1901	Additional Surgery	A070504	Premature Discharge of Battery	19/06/2023	19/06/2023
									A090805	Incorrect Measurement	19/06/2023	19/06/2023
					CT3002	Implantable synthetic polymer	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	19/06/2023	19/06/2023
									A090805	Incorrect Measurement	19/06/2023	19/06/2023
							F1901	Additional Surgery	A070504	Premature Discharge of Battery	19/06/2023	19/06/2023
									A090805	Incorrect Measurement	19/06/2023	19/06/2023
					CT3006	Implantable inorganic material	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	19/06/2023	19/06/2023
									A090805	Incorrect Measurement	19/06/2023	19/06/2023
							F1901	Additional Surgery	A070504	Premature Discharge of Battery	19/06/2023	19/06/2023
									A090805	Incorrect Measurement	19/06/2023	19/06/2023
					CT314	Electrically powered	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	19/06/2023	19/06/2023
									A090805	Incorrect Measurement	19/06/2023	19/06/2023
							F1901	Additional Surgery	A070504	Premature Discharge of Battery	19/06/2023	19/06/2023
									A090805	Incorrect Measurement	19/06/2023	19/06/2023
					CT321	Long-term surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	19/06/2023	19/06/2023
									A090805	Incorrect Measurement	19/06/2023	19/06/2023
							F1901	Additional Surgery	A070504	Premature Discharge of Battery	19/06/2023	19/06/2023
									A090805	Incorrect Measurement	19/06/2023	19/06/2023
					CT449	Generators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	19/06/2023	19/06/2023
									A090805	Incorrect Measurement	19/06/2023	19/06/2023
							F1901	Additional Surgery	A070504	Premature Discharge of Battery	19/06/2023	19/06/2023
									A090805	Incorrect Measurement	19/06/2023	19/06/2023
					CT981	Single use	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	19/06/2023	19/06/2023
									A090805	Incorrect Measurement	19/06/2023	19/06/2023

							F1901	Additional Surgery	A070504	Premature Discharge of Battery	19/06/2023	19/06/2023
									A090805	Incorrect Measurement	19/06/2023	19/06/2023
					CT983	Surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	19/06/2023	19/06/2023
									A090805	Incorrect Measurement	19/06/2023	19/06/2023
							F1901	Additional Surgery	A070504	Premature Discharge of Battery	19/06/2023	19/06/2023
									A090805	Incorrect Measurement	19/06/2023	19/06/2023
13154	2023		47265	Dual-chamber implantable pacemaker, rate-responsive	CT141	Electrodes/Leads and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	23/02/2023	23/02/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	23/02/2023	23/02/2023
					CT145	Cardiovascular devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	23/02/2023	23/02/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	23/02/2023	23/02/2023
					CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	23/02/2023	23/02/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	23/02/2023	23/02/2023
					CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	23/02/2023	23/02/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	23/02/2023	23/02/2023
					CT233	Surgical	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	23/02/2023	23/02/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	23/02/2023	23/02/2023
					CT2406	Implantable	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	23/02/2023	23/02/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	23/02/2023	23/02/2023
					CT2407	Active implantable	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	23/02/2023	23/02/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	23/02/2023	23/02/2023
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	23/02/2023	23/02/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	23/02/2023	23/02/2023
					CT2727	Active implant devices/ancillaries	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	23/02/2023	23/02/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	23/02/2023	23/02/2023
					CT2750	Implantable cardiac electrical stimulators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	23/02/2023	23/02/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	23/02/2023	23/02/2023
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	23/02/2023	23/02/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	23/02/2023	23/02/2023
					CT2985	Active	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	23/02/2023	23/02/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	23/02/2023	23/02/2023
					CT3001	Implantable metal	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	23/02/2023	23/02/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	23/02/2023	23/02/2023
					CT3002	Implantable synthetic polymer	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	23/02/2023	23/02/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	23/02/2023	23/02/2023
					CT3006	Implantable inorganic material	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	23/02/2023	23/02/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	23/02/2023	23/02/2023
					CT314	Electrically powered	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	23/02/2023	23/02/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	23/02/2023	23/02/2023
					CT321	Long-term surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	23/02/2023	23/02/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	23/02/2023	23/02/2023
					CT449	Generators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	23/02/2023	23/02/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	23/02/2023	23/02/2023
					CT981	Single use	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	23/02/2023	23/02/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	23/02/2023	23/02/2023
					CT983	Surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	23/02/2023	23/02/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	23/02/2023	23/02/2023
13155	2023	All other reportable incidents	47265	Dual-chamber implantable pacemaker, rate-responsive	CT141	Electrodes/Leads and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070909	Oversensing		
							F26	No Health Consequences or Impact	A070909	Oversensing		
					CT145	Cardiovascular devices	E2403	No Clinical Signs, Symptoms or Conditions	A070909	Oversensing		
							F26	No Health Consequences or Impact	A070909	Oversensing		
					CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070909	Oversensing		
							F26	No Health Consequences or Impact	A070909	Oversensing		
					CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A070909	Oversensing		
							F26	No Health Consequences or Impact	A070909	Oversensing		
					CT233	Surgical	E2403	No Clinical Signs, Symptoms or Conditions	A070909	Oversensing		
							F26	No Health Consequences or Impact	A070909	Oversensing		
					CT2406	Implantable	E2403	No Clinical Signs, Symptoms or Conditions	A070909	Oversensing		
							F26	No Health Consequences or Impact	A070909	Oversensing		
					CT2407	Active implantable	E2403	No Clinical Signs, Symptoms or Conditions	A070909	Oversensing		
							F26	No Health Consequences or Impact	A070909	Oversensing		
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A070909	Oversensing		
							F26	No Health Consequences or Impact	A070909	Oversensing		
					CT2727	Active implant devices/ancillaries	E2403	No Clinical Signs, Symptoms or Conditions	A070909	Oversensing		
							F26	No Health Consequences or Impact	A070909	Oversensing		
					CT2750	Implantable cardiac electrical stimulators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070909	Oversensing		
							F26	No Health Consequences or Impact	A070909	Oversensing		
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A070909	Oversensing		
							F26	No Health Consequences or Impact	A070909	Oversensing		
					CT2985	Active	E2403	No Clinical Signs, Symptoms or Conditions	A070909	Oversensing		
							F26	No Health Consequences or Impact	A070909	Oversensing		
					CT3001	Implantable metal	E2403	No Clinical Signs, Symptoms or Conditions	A070909	Oversensing		
							F26	No Health Consequences or Impact	A070909	Oversensing		
					CT3002	Implantable synthetic polymer	E2403	No Clinical Signs, Symptoms or Conditions	A070909	Oversensing		
							F26	No Health Consequences or Impact	A070909	Oversensing		
					CT3006	Implantable inorganic material	E2403	No Clinical Signs, Symptoms or Conditions	A070909	Oversensing		
							F26	No Health Consequences or Impact	A070909	Oversensing		
					CT314	Electrically powered	E2403	No Clinical Signs, Symptoms or Conditions	A070909	Oversensing		
							F26	No Health Consequences or Impact	A070909	Oversensing		
					CT321	Long-term surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A070909	Oversensing		
							F26	No Health Consequences or Impact	A070909	Oversensing		
					CT449	Generators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070909	Oversensing		
							F26	No Health Consequences or Impact	A070909	Oversensing		
					CT981	Single use	E2403	No Clinical Signs, Symptoms or Conditions	A070909	Oversensing		
							F26	No Health Consequences or Impact	A070909	Oversensing		
					CT983	Surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A070909	Oversensing		
							F26	No Health Consequences or Impact	A070909	Oversensing		
13156	2023		47265	Dual-chamber implantable pacemaker, rate-responsive	CT141	Electrodes/Leads and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A0203	Defective Device	01/12/2022	01/12/2022
									A072201	High Impedance	01/12/2022	01/12/2022

							F1901	Additional Surgery	A0203	Defective Device	01/12/2022	01/12/2022
									A072201	High Impedance	01/12/2022	01/12/2022
					CT145	Cardiovascular devices	E2403	No Clinical Signs, Symptoms or Conditions	A0203	Defective Device	01/12/2022	01/12/2022
									A072201	High Impedance	01/12/2022	01/12/2022
							F1901	Additional Surgery	A0203	Defective Device	01/12/2022	01/12/2022
									A072201	High Impedance	01/12/2022	01/12/2022
					CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A0203	Defective Device	01/12/2022	01/12/2022
									A072201	High Impedance	01/12/2022	01/12/2022
							F1901	Additional Surgery	A0203	Defective Device	01/12/2022	01/12/2022
									A072201	High Impedance	01/12/2022	01/12/2022
					CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A0203	Defective Device	01/12/2022	01/12/2022
									A072201	High Impedance	01/12/2022	01/12/2022
							F1901	Additional Surgery	A0203	Defective Device	01/12/2022	01/12/2022
									A072201	High Impedance	01/12/2022	01/12/2022
					CT233	Surgical	E2403	No Clinical Signs, Symptoms or Conditions	A0203	Defective Device	01/12/2022	01/12/2022
									A072201	High Impedance	01/12/2022	01/12/2022
							F1901	Additional Surgery	A0203	Defective Device	01/12/2022	01/12/2022
									A072201	High Impedance	01/12/2022	01/12/2022
					CT2406	Implantable	E2403	No Clinical Signs, Symptoms or Conditions	A0203	Defective Device	01/12/2022	01/12/2022
									A072201	High Impedance	01/12/2022	01/12/2022
							F1901	Additional Surgery	A0203	Defective Device	01/12/2022	01/12/2022
									A072201	High Impedance	01/12/2022	01/12/2022
					CT2407	Active implantable	E2403	No Clinical Signs, Symptoms or Conditions	A0203	Defective Device	01/12/2022	01/12/2022
									A072201	High Impedance	01/12/2022	01/12/2022
							F1901	Additional Surgery	A0203	Defective Device	01/12/2022	01/12/2022
									A072201	High Impedance	01/12/2022	01/12/2022
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A0203	Defective Device	01/12/2022	01/12/2022
									A072201	High Impedance	01/12/2022	01/12/2022
							F1901	Additional Surgery	A0203	Defective Device	01/12/2022	01/12/2022
									A072201	High Impedance	01/12/2022	01/12/2022
					CT2727	Active implant devices/ancillaries	E2403	No Clinical Signs, Symptoms or Conditions	A0203	Defective Device	01/12/2022	01/12/2022
									A072201	High Impedance	01/12/2022	01/12/2022
							F1901	Additional Surgery	A0203	Defective Device	01/12/2022	01/12/2022
									A072201	High Impedance	01/12/2022	01/12/2022
					CT2750	Implantable cardiac electrical stimulators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A0203	Defective Device	01/12/2022	01/12/2022
									A072201	High Impedance	01/12/2022	01/12/2022
							F1901	Additional Surgery	A0203	Defective Device	01/12/2022	01/12/2022
									A072201	High Impedance	01/12/2022	01/12/2022
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A0203	Defective Device	01/12/2022	01/12/2022
									A072201	High Impedance	01/12/2022	01/12/2022
							F1901	Additional Surgery	A0203	Defective Device	01/12/2022	01/12/2022
									A072201	High Impedance	01/12/2022	01/12/2022
					CT2985	Active	E2403	No Clinical Signs, Symptoms or Conditions	A0203	Defective Device	01/12/2022	01/12/2022
									A072201	High Impedance	01/12/2022	01/12/2022
							F1901	Additional Surgery	A0203	Defective Device	01/12/2022	01/12/2022
									A072201	High Impedance	01/12/2022	01/12/2022
					CT3001	Implantable metal	E2403	No Clinical Signs, Symptoms or Conditions	A0203	Defective Device	01/12/2022	01/12/2022
									A072201	High Impedance	01/12/2022	01/12/2022
							F1901	Additional Surgery	A0203	Defective Device	01/12/2022	01/12/2022
									A072201	High Impedance	01/12/2022	01/12/2022
					CT3002	Implantable synthetic polymer	E2403	No Clinical Signs, Symptoms or Conditions	A0203	Defective Device	01/12/2022	01/12/2022
									A072201	High Impedance	01/12/2022	01/12/2022
							F1901	Additional Surgery	A0203	Defective Device	01/12/2022	01/12/2022
									A072201	High Impedance	01/12/2022	01/12/2022
					CT3006	Implantable inorganic material	E2403	No Clinical Signs, Symptoms or Conditions	A0203	Defective Device	01/12/2022	01/12/2022
									A072201	High Impedance	01/12/2022	01/12/2022
							F1901	Additional Surgery	A0203	Defective Device	01/12/2022	01/12/2022
									A072201	High Impedance	01/12/2022	01/12/2022
					CT314	Electrically powered	E2403	No Clinical Signs, Symptoms or Conditions	A0203	Defective Device	01/12/2022	01/12/2022
									A072201	High Impedance	01/12/2022	01/12/2022
							F1901	Additional Surgery	A0203	Defective Device	01/12/2022	01/12/2022
									A072201	High Impedance	01/12/2022	01/12/2022
					CT321	Long-term surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A0203	Defective Device	01/12/2022	01/12/2022
									A072201	High Impedance	01/12/2022	01/12/2022
							F1901	Additional Surgery	A0203	Defective Device	01/12/2022	01/12/2022
									A072201	High Impedance	01/12/2022	01/12/2022
					CT449	Generators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A0203	Defective Device	01/12/2022	01/12/2022
									A072201	High Impedance	01/12/2022	01/12/2022
							F1901	Additional Surgery	A0203	Defective Device	01/12/2022	01/12/2022
									A072201	High Impedance	01/12/2022	01/12/2022
					CT981	Single use	E2403	No Clinical Signs, Symptoms or Conditions	A0203	Defective Device	01/12/2022	01/12/2022
									A072201	High Impedance	01/12/2022	01/12/2022
							F1901	Additional Surgery	A0203	Defective Device	01/12/2022	01/12/2022
									A072201	High Impedance	01/12/2022	01/12/2022
					CT983	Surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A0203	Defective Device	01/12/2022	01/12/2022
									A072201	High Impedance	01/12/2022	01/12/2022
							F1901	Additional Surgery	A0203	Defective Device	01/12/2022	01/12/2022
									A072201	High Impedance	01/12/2022	01/12/2022
13157	2023	All other reportable incidents	47265	Dual-chamber implantable pacemaker, rate-responsive	CT141	Electrodes/Leads and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A0203	Defective Device	06/01/2023	06/01/2023
									A070909	Oversensing	06/01/2023	06/01/2023
							F1901	Additional Surgery	A0203	Defective Device	06/01/2023	06/01/2023
									A070909	Oversensing	06/01/2023	06/01/2023
					CT145	Cardiovascular devices	E2403	No Clinical Signs, Symptoms or Conditions	A0203	Defective Device	06/01/2023	06/01/2023
									A070909	Oversensing	06/01/2023	06/01/2023
							F1901	Additional Surgery	A0203	Defective Device	06/01/2023	06/01/2023
									A070909	Oversensing	06/01/2023	06/01/2023
					CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A0203	Defective Device	06/01/2023	06/01/2023
									A070909	Oversensing	06/01/2023	06/01/2023

							F1901	Additional Surgery	A0203	Defective Device	06/01/2023	06/01/2023
									A070909	Oversensing	06/01/2023	06/01/2023
					CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A0203	Defective Device	06/01/2023	06/01/2023
									A070909	Oversensing	06/01/2023	06/01/2023
							F1901	Additional Surgery	A0203	Defective Device	06/01/2023	06/01/2023
									A070909	Oversensing	06/01/2023	06/01/2023
					CT233	Surgical	E2403	No Clinical Signs, Symptoms or Conditions	A0203	Defective Device	06/01/2023	06/01/2023
									A070909	Oversensing	06/01/2023	06/01/2023
							F1901	Additional Surgery	A0203	Defective Device	06/01/2023	06/01/2023
									A070909	Oversensing	06/01/2023	06/01/2023
					CT2406	Implantable	E2403	No Clinical Signs, Symptoms or Conditions	A0203	Defective Device	06/01/2023	06/01/2023
									A070909	Oversensing	06/01/2023	06/01/2023
							F1901	Additional Surgery	A0203	Defective Device	06/01/2023	06/01/2023
									A070909	Oversensing	06/01/2023	06/01/2023
					CT2407	Active implantable	E2403	No Clinical Signs, Symptoms or Conditions	A0203	Defective Device	06/01/2023	06/01/2023
									A070909	Oversensing	06/01/2023	06/01/2023
							F1901	Additional Surgery	A0203	Defective Device	06/01/2023	06/01/2023
									A070909	Oversensing	06/01/2023	06/01/2023
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A0203	Defective Device	06/01/2023	06/01/2023
									A070909	Oversensing	06/01/2023	06/01/2023
							F1901	Additional Surgery	A0203	Defective Device	06/01/2023	06/01/2023
									A070909	Oversensing	06/01/2023	06/01/2023
					CT2727	Active implant devices/ancillaries	E2403	No Clinical Signs, Symptoms or Conditions	A0203	Defective Device	06/01/2023	06/01/2023
									A070909	Oversensing	06/01/2023	06/01/2023
							F1901	Additional Surgery	A0203	Defective Device	06/01/2023	06/01/2023
									A070909	Oversensing	06/01/2023	06/01/2023
					CT2750	Implantable cardiac electrical stimulators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A0203	Defective Device	06/01/2023	06/01/2023
									A070909	Oversensing	06/01/2023	06/01/2023
							F1901	Additional Surgery	A0203	Defective Device	06/01/2023	06/01/2023
									A070909	Oversensing	06/01/2023	06/01/2023
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A0203	Defective Device	06/01/2023	06/01/2023
									A070909	Oversensing	06/01/2023	06/01/2023
							F1901	Additional Surgery	A0203	Defective Device	06/01/2023	06/01/2023
									A070909	Oversensing	06/01/2023	06/01/2023
					CT2985	Active	E2403	No Clinical Signs, Symptoms or Conditions	A0203	Defective Device	06/01/2023	06/01/2023
									A070909	Oversensing	06/01/2023	06/01/2023
							F1901	Additional Surgery	A0203	Defective Device	06/01/2023	06/01/2023
									A070909	Oversensing	06/01/2023	06/01/2023
					CT3001	Implantable metal	E2403	No Clinical Signs, Symptoms or Conditions	A0203	Defective Device	06/01/2023	06/01/2023
									A070909	Oversensing	06/01/2023	06/01/2023
							F1901	Additional Surgery	A0203	Defective Device	06/01/2023	06/01/2023
									A070909	Oversensing	06/01/2023	06/01/2023
					CT3002	Implantable synthetic polymer	E2403	No Clinical Signs, Symptoms or Conditions	A0203	Defective Device	06/01/2023	06/01/2023
									A070909	Oversensing	06/01/2023	06/01/2023
							F1901	Additional Surgery	A0203	Defective Device	06/01/2023	06/01/2023
									A070909	Oversensing	06/01/2023	06/01/2023
					CT3006	Implantable inorganic material	E2403	No Clinical Signs, Symptoms or Conditions	A0203	Defective Device	06/01/2023	06/01/2023
									A070909	Oversensing	06/01/2023	06/01/2023
							F1901	Additional Surgery	A0203	Defective Device	06/01/2023	06/01/2023
									A070909	Oversensing	06/01/2023	06/01/2023
					CT314	Electrically powered	E2403	No Clinical Signs, Symptoms or Conditions	A0203	Defective Device	06/01/2023	06/01/2023
									A070909	Oversensing	06/01/2023	06/01/2023
							F1901	Additional Surgery	A0203	Defective Device	06/01/2023	06/01/2023
									A070909	Oversensing	06/01/2023	06/01/2023
					CT321	Long-term surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A0203	Defective Device	06/01/2023	06/01/2023
									A070909	Oversensing	06/01/2023	06/01/2023
							F1901	Additional Surgery	A0203	Defective Device	06/01/2023	06/01/2023
									A070909	Oversensing	06/01/2023	06/01/2023
					CT449	Generators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A0203	Defective Device	06/01/2023	06/01/2023
									A070909	Oversensing	06/01/2023	06/01/2023
							F1901	Additional Surgery	A0203	Defective Device	06/01/2023	06/01/2023
									A070909	Oversensing	06/01/2023	06/01/2023
					CT981	Single use	E2403	No Clinical Signs, Symptoms or Conditions	A0203	Defective Device	06/01/2023	06/01/2023
									A070909	Oversensing	06/01/2023	06/01/2023
							F1901	Additional Surgery	A0203	Defective Device	06/01/2023	06/01/2023
									A070909	Oversensing	06/01/2023	06/01/2023
					CT983	Surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A0203	Defective Device	06/01/2023	06/01/2023
									A070909	Oversensing	06/01/2023	06/01/2023
							F1901	Additional Surgery	A0203	Defective Device	06/01/2023	06/01/2023
									A070909	Oversensing	06/01/2023	06/01/2023
13158	2023	All other reportable incidents	35223	Endocardial/interventricular septal pacing lead	CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070909	Oversensing		
							F1901	Additional Surgery	A070909	Oversensing		
13159	2023	All other reportable incidents	35223	Endocardial/interventricular septal pacing lead	CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A0401	Break		
									A072201	High Impedance		
							F1901	Additional Surgery	A0401	Break		
									A072201	High Impedance		
13160	2023		47267	Single-chamber implantable pacemaker, rate-responsive	CT141	Electrodes/Leads and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A072201	High Impedance	01/03/2023	01/03/2023
							F1901	Additional Surgery	A072201	High Impedance	01/03/2023	01/03/2023
					CT145	Cardiovascular devices	E2403	No Clinical Signs, Symptoms or Conditions	A072201	High Impedance	01/03/2023	01/03/2023
							F1901	Additional Surgery	A072201	High Impedance	01/03/2023	01/03/2023
					CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A072201	High Impedance	01/03/2023	01/03/2023
							F1901	Additional Surgery	A072201	High Impedance	01/03/2023	01/03/2023
					CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A072201	High Impedance	01/03/2023	01/03/2023
							F1901	Additional Surgery	A072201	High Impedance	01/03/2023	01/03/2023
					CT233	Surgical	E2403	No Clinical Signs, Symptoms or Conditions	A072201	High Impedance	01/03/2023	01/03/2023
							F1901	Additional Surgery	A072201	High Impedance	01/03/2023	01/03/2023
					CT2406	Implantable	E2403	No Clinical Signs, Symptoms or Conditions	A072201	High Impedance	01/03/2023	01/03/2023
							F1901	Additional Surgery	A072201	High Impedance	01/03/2023	01/03/2023

					CT2407	Active implantable	E2403	No Clinical Signs, Symptoms or Conditions	A072201	High Impedance	01/03/2023	01/03/2023
							F1901	Additional Surgery	A072201	High Impedance	01/03/2023	01/03/2023
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A072201	High Impedance	01/03/2023	01/03/2023
							F1901	Additional Surgery	A072201	High Impedance	01/03/2023	01/03/2023
					CT2727	Active implant devices/ancillaries	E2403	No Clinical Signs, Symptoms or Conditions	A072201	High Impedance	01/03/2023	01/03/2023
							F1901	Additional Surgery	A072201	High Impedance	01/03/2023	01/03/2023
					CT2750	Implantable cardiac electrical stimulators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A072201	High Impedance	01/03/2023	01/03/2023
							F1901	Additional Surgery	A072201	High Impedance	01/03/2023	01/03/2023
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A072201	High Impedance	01/03/2023	01/03/2023
							F1901	Additional Surgery	A072201	High Impedance	01/03/2023	01/03/2023
					CT2985	Active	E2403	No Clinical Signs, Symptoms or Conditions	A072201	High Impedance	01/03/2023	01/03/2023
							F1901	Additional Surgery	A072201	High Impedance	01/03/2023	01/03/2023
					CT3001	Implantable metal	E2403	No Clinical Signs, Symptoms or Conditions	A072201	High Impedance	01/03/2023	01/03/2023
							F1901	Additional Surgery	A072201	High Impedance	01/03/2023	01/03/2023
					CT3002	Implantable synthetic polymer	E2403	No Clinical Signs, Symptoms or Conditions	A072201	High Impedance	01/03/2023	01/03/2023
							F1901	Additional Surgery	A072201	High Impedance	01/03/2023	01/03/2023
					CT3006	Implantable inorganic material	E2403	No Clinical Signs, Symptoms or Conditions	A072201	High Impedance	01/03/2023	01/03/2023
							F1901	Additional Surgery	A072201	High Impedance	01/03/2023	01/03/2023
					CT314	Electrically powered	E2403	No Clinical Signs, Symptoms or Conditions	A072201	High Impedance	01/03/2023	01/03/2023
							F1901	Additional Surgery	A072201	High Impedance	01/03/2023	01/03/2023
					CT321	Long-term surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A072201	High Impedance	01/03/2023	01/03/2023
							F1901	Additional Surgery	A072201	High Impedance	01/03/2023	01/03/2023
					CT449	Generators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A072201	High Impedance	01/03/2023	01/03/2023
							F1901	Additional Surgery	A072201	High Impedance	01/03/2023	01/03/2023
					CT981	Single use	E2403	No Clinical Signs, Symptoms or Conditions	A072201	High Impedance	01/03/2023	01/03/2023
							F1901	Additional Surgery	A072201	High Impedance	01/03/2023	01/03/2023
					CT983	Surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A072201	High Impedance	01/03/2023	01/03/2023
							F1901	Additional Surgery	A072201	High Impedance	01/03/2023	01/03/2023
13161	2023		47270	Cardiac resynchronization therapy implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A070909	Oversensing	01/03/2023	31/03/2023
							F1901	Additional Surgery	A070909	Oversensing	01/03/2023	31/03/2023
13162	2023	All other reportable incidents	47265	Dual-chamber implantable pacemaker, rate-responsive	CT141	Electrodes/Leads and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070909	Oversensing	01/03/2023	01/03/2023
							F1901	Additional Surgery	A070909	Oversensing	01/03/2023	01/03/2023
					CT145	Cardiovascular devices	E2403	No Clinical Signs, Symptoms or Conditions	A070909	Oversensing	01/03/2023	01/03/2023
							F1901	Additional Surgery	A070909	Oversensing	01/03/2023	01/03/2023
					CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070909	Oversensing	01/03/2023	01/03/2023
							F1901	Additional Surgery	A070909	Oversensing	01/03/2023	01/03/2023
					CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A070909	Oversensing	01/03/2023	01/03/2023
							F1901	Additional Surgery	A070909	Oversensing	01/03/2023	01/03/2023
					CT233	Surgical	E2403	No Clinical Signs, Symptoms or Conditions	A070909	Oversensing	01/03/2023	01/03/2023
							F1901	Additional Surgery	A070909	Oversensing	01/03/2023	01/03/2023
					CT2406	Implantable	E2403	No Clinical Signs, Symptoms or Conditions	A070909	Oversensing	01/03/2023	01/03/2023
							F1901	Additional Surgery	A070909	Oversensing	01/03/2023	01/03/2023
					CT2407	Active implantable	E2403	No Clinical Signs, Symptoms or Conditions	A070909	Oversensing	01/03/2023	01/03/2023
							F1901	Additional Surgery	A070909	Oversensing	01/03/2023	01/03/2023
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A070909	Oversensing	01/03/2023	01/03/2023
							F1901	Additional Surgery	A070909	Oversensing	01/03/2023	01/03/2023
					CT2727	Active implant devices/ancillaries	E2403	No Clinical Signs, Symptoms or Conditions	A070909	Oversensing	01/03/2023	01/03/2023
							F1901	Additional Surgery	A070909	Oversensing	01/03/2023	01/03/2023
					CT2750	Implantable cardiac electrical stimulators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070909	Oversensing	01/03/2023	01/03/2023
							F1901	Additional Surgery	A070909	Oversensing	01/03/2023	01/03/2023
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A070909	Oversensing	01/03/2023	01/03/2023
							F1901	Additional Surgery	A070909	Oversensing	01/03/2023	01/03/2023
					CT2985	Active	E2403	No Clinical Signs, Symptoms or Conditions	A070909	Oversensing	01/03/2023	01/03/2023
							F1901	Additional Surgery	A070909	Oversensing	01/03/2023	01/03/2023
					CT3001	Implantable metal	E2403	No Clinical Signs, Symptoms or Conditions	A070909	Oversensing	01/03/2023	01/03/2023
							F1901	Additional Surgery	A070909	Oversensing	01/03/2023	01/03/2023
					CT3002	Implantable synthetic polymer	E2403	No Clinical Signs, Symptoms or Conditions	A070909	Oversensing	01/03/2023	01/03/2023
							F1901	Additional Surgery	A070909	Oversensing	01/03/2023	01/03/2023
					CT3006	Implantable inorganic material	E2403	No Clinical Signs, Symptoms or Conditions	A070909	Oversensing	01/03/2023	01/03/2023
							F1901	Additional Surgery	A070909	Oversensing	01/03/2023	01/03/2023
					CT314	Electrically powered	E2403	No Clinical Signs, Symptoms or Conditions	A070909	Oversensing	01/03/2023	01/03/2023
							F1901	Additional Surgery	A070909	Oversensing	01/03/2023	01/03/2023
					CT321	Long-term surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A070909	Oversensing	01/03/2023	01/03/2023
							F1901	Additional Surgery	A070909	Oversensing	01/03/2023	01/03/2023
					CT449	Generators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070909	Oversensing	01/03/2023	01/03/2023
							F1901	Additional Surgery	A070909	Oversensing	01/03/2023	01/03/2023
					CT981	Single use	E2403	No Clinical Signs, Symptoms or Conditions	A070909	Oversensing	01/03/2023	01/03/2023
							F1901	Additional Surgery	A070909	Oversensing	01/03/2023	01/03/2023
					CT983	Surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A070909	Oversensing	01/03/2023	01/03/2023
							F1901	Additional Surgery	A070909	Oversensing	01/03/2023	01/03/2023
13163	2023	All other reportable incidents	35223	Endocardial/interventricular septal pacing lead	CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070909	Oversensing		
							F1901	Additional Surgery	A070909	Oversensing		
13164	2023	All other reportable incidents	35852	Single-chamber implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
									A0908	Incorrect, Inadequate or Imprecise Result or Readings		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
									A0908	Incorrect, Inadequate or Imprecise Result or Readings		
13165	2023	All other reportable incidents	35852	Single-chamber implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E1705	Burning Sensation	A0709	Device Sensing Problem		
									A070909	Oversensing		
									A071302	Inappropriate/Inadequate Shock/Stimulation		
									A12	Connection Problem		
							E2104	Electric Shock	A0709	Device Sensing Problem		
									A070909	Oversensing		
									A071302	Inappropriate/Inadequate Shock/Stimulation		
									A12	Connection Problem		
							F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A0709	Device Sensing Problem		
									A070909	Oversensing		

								A071302	Inappropriate/Inadequate Shock/Stimulation		
								A12	Connection Problem		
						F2203	Imaging Required	A0709	Device Sensing Problem		
								A070909	Oversensing		
								A071302	Inappropriate/Inadequate Shock/Stimulation		
								A12	Connection Problem		
13166	2023		47265	Dual-chamber implantable pacemaker, rate-responsive	CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A0203	Defective Device		
									A1304	Telemetry Discrepancy		
							F1901	Additional Surgery	A0203	Defective Device		
									A1304	Telemetry Discrepancy		
							F2301	Additional Device Required	A0203	Defective Device		
									A1304	Telemetry Discrepancy		
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A0203	Defective Device		
									A1304	Telemetry Discrepancy		
							F1901	Additional Surgery	A0203	Defective Device		
									A1304	Telemetry Discrepancy		
							F2301	Additional Device Required	A0203	Defective Device		
									A1304	Telemetry Discrepancy		
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A0203	Defective Device		
									A1304	Telemetry Discrepancy		
							F1901	Additional Surgery	A0203	Defective Device		
									A1304	Telemetry Discrepancy		
							F2301	Additional Device Required	A0203	Defective Device		
									A1304	Telemetry Discrepancy		
13167	2023		47270	Cardiac resynchronization therapy implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A160602	Premature End-of-Life Indicator		
							F26	No Health Consequences or Impact	A160602	Premature End-of-Life Indicator		
13168	2023	All other reportable incidents	47265	Dual-chamber implantable pacemaker, rate-responsive	CT141	Electrodes/Leads and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT145	Cardiovascular devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT233	Surgical	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT2406	Implantable	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT2407	Active implantable	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT2727	Active implant devices/ancillaries	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT2750	Implantable cardiac electrical stimulators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT2985	Active	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT3001	Implantable metal	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT3002	Implantable synthetic polymer	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT3006	Implantable inorganic material	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT314	Electrically powered	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT321	Long-term surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT449	Generators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT981	Single use	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT983	Surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
13169	2023		35852	Single-chamber implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E0601	Arrhythmia	A071201	Failure to Convert Rhythm		
							F12	Serious Injury/ Illness/ Impairment	A071201	Failure to Convert Rhythm		
13170	2023		47265	Dual-chamber implantable pacemaker, rate-responsive	CT141	Electrodes/Leads and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	11/05/2023	11/05/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	11/05/2023	11/05/2023
					CT145	Cardiovascular devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	11/05/2023	11/05/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	11/05/2023	11/05/2023
					CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	11/05/2023	11/05/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	11/05/2023	11/05/2023
					CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	11/05/2023	11/05/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	11/05/2023	11/05/2023
					CT233	Surgical	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	11/05/2023	11/05/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	11/05/2023	11/05/2023
					CT2406	Implantable	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	11/05/2023	11/05/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	11/05/2023	11/05/2023
					CT2407	Active implantable	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	11/05/2023	11/05/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	11/05/2023	11/05/2023
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	11/05/2023	11/05/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	11/05/2023	11/05/2023
					CT2727	Active implant devices/ancillaries	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	11/05/2023	11/05/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	11/05/2023	11/05/2023
					CT2750	Implantable cardiac electrical stimulators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	11/05/2023	11/05/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	11/05/2023	11/05/2023

					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	11/05/2023	11/05/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	11/05/2023	11/05/2023
					CT2985	Active	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	11/05/2023	11/05/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	11/05/2023	11/05/2023
					CT3001	Implantable metal	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	11/05/2023	11/05/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	11/05/2023	11/05/2023
					CT3002	Implantable synthetic polymer	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	11/05/2023	11/05/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	11/05/2023	11/05/2023
					CT3006	Implantable inorganic material	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	11/05/2023	11/05/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	11/05/2023	11/05/2023
					CT314	Electrically powered	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	11/05/2023	11/05/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	11/05/2023	11/05/2023
					CT321	Long-term surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	11/05/2023	11/05/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	11/05/2023	11/05/2023
					CT449	Generators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	11/05/2023	11/05/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	11/05/2023	11/05/2023
					CT981	Single use	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	11/05/2023	11/05/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	11/05/2023	11/05/2023
					CT983	Surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	11/05/2023	11/05/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	11/05/2023	11/05/2023
13171	2023		47265	Dual-chamber implantable pacemaker, rate-responsive	CT141	Electrodes/Leads and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	11/05/2023	11/05/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	11/05/2023	11/05/2023
					CT145	Cardiovascular devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	11/05/2023	11/05/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	11/05/2023	11/05/2023
					CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	11/05/2023	11/05/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	11/05/2023	11/05/2023
					CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	11/05/2023	11/05/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	11/05/2023	11/05/2023
					CT233	Surgical	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	11/05/2023	11/05/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	11/05/2023	11/05/2023
					CT2406	Implantable	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	11/05/2023	11/05/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	11/05/2023	11/05/2023
					CT2407	Active implantable	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	11/05/2023	11/05/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	11/05/2023	11/05/2023
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	11/05/2023	11/05/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	11/05/2023	11/05/2023
					CT2727	Active implant devices/ancillaries	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	11/05/2023	11/05/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	11/05/2023	11/05/2023
					CT2750	Implantable cardiac electrical stimulators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	11/05/2023	11/05/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	11/05/2023	11/05/2023
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	11/05/2023	11/05/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	11/05/2023	11/05/2023
					CT2985	Active	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	11/05/2023	11/05/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	11/05/2023	11/05/2023
					CT3001	Implantable metal	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	11/05/2023	11/05/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	11/05/2023	11/05/2023
					CT3002	Implantable synthetic polymer	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	11/05/2023	11/05/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	11/05/2023	11/05/2023
					CT3006	Implantable inorganic material	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	11/05/2023	11/05/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	11/05/2023	11/05/2023
					CT314	Electrically powered	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	11/05/2023	11/05/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	11/05/2023	11/05/2023
					CT321	Long-term surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	11/05/2023	11/05/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	11/05/2023	11/05/2023
					CT449	Generators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	11/05/2023	11/05/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	11/05/2023	11/05/2023
					CT981	Single use	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	11/05/2023	11/05/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	11/05/2023	11/05/2023
					CT983	Surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	11/05/2023	11/05/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	11/05/2023	11/05/2023
13172	2023	All other reportable incidents	17882	Physiologic-monitoring defibrillation system	CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A11	Computer Software Problem		
							F27	Problem Identified During Non-Clinical Procedure	A11	Computer Software Problem		
13173	2023	All other reportable incidents	17882	Physiologic-monitoring defibrillation system	CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A150105	Key or Button Unresponsive/not Working		
							F27	Problem Identified During Non-Clinical Procedure	A150105	Key or Button Unresponsive/not Working		
13174	2023		47263	Cardiac resynchronization therapy implantable pacemaker	CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
13175	2023	All other reportable incidents	60789	Intracardiac pacemaker	CT130	Catheters and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A05	Mechanical Problem		
									A0504	Leak/Splash		
							F26	No Health Consequences or Impact	A05	Mechanical Problem		
									A0504	Leak/Splash		
					CT145	Cardiovascular devices	E2403	No Clinical Signs, Symptoms or Conditions	A05	Mechanical Problem		
									A0504	Leak/Splash		
							F26	No Health Consequences or Impact	A05	Mechanical Problem		
									A0504	Leak/Splash		
					CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A05	Mechanical Problem		
									A0504	Leak/Splash		
							F26	No Health Consequences or Impact	A05	Mechanical Problem		
									A0504	Leak/Splash		
					CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A05	Mechanical Problem		
									A0504	Leak/Splash		
							F26	No Health Consequences or Impact	A05	Mechanical Problem		
									A0504	Leak/Splash		

					CT233	Surgical	E2403	No Clinical Signs, Symptoms or Conditions	A05	Mechanical Problem		
									A0504	Leak/Splash		
							F26	No Health Consequences or Impact	A05	Mechanical Problem		
									A0504	Leak/Splash		
					CT2406	Implantable	E2403	No Clinical Signs, Symptoms or Conditions	A05	Mechanical Problem		
									A0504	Leak/Splash		
							F26	No Health Consequences or Impact	A05	Mechanical Problem		
									A0504	Leak/Splash		
					CT2407	Active implantable	E2403	No Clinical Signs, Symptoms or Conditions	A05	Mechanical Problem		
									A0504	Leak/Splash		
							F26	No Health Consequences or Impact	A05	Mechanical Problem		
									A0504	Leak/Splash		
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A05	Mechanical Problem		
									A0504	Leak/Splash		
							F26	No Health Consequences or Impact	A05	Mechanical Problem		
									A0504	Leak/Splash		
					CT2727	Active implant devices/ancillaries	E2403	No Clinical Signs, Symptoms or Conditions	A05	Mechanical Problem		
									A0504	Leak/Splash		
							F26	No Health Consequences or Impact	A05	Mechanical Problem		
									A0504	Leak/Splash		
					CT2750	Implantable cardiac electrical stimulators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A05	Mechanical Problem		
									A0504	Leak/Splash		
							F26	No Health Consequences or Impact	A05	Mechanical Problem		
									A0504	Leak/Splash		
					CT2818	Catheterization/Cannulation/Fluid path devices	E2403	No Clinical Signs, Symptoms or Conditions	A05	Mechanical Problem		
									A0504	Leak/Splash		
							F26	No Health Consequences or Impact	A05	Mechanical Problem		
									A0504	Leak/Splash		
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A05	Mechanical Problem		
									A0504	Leak/Splash		
							F26	No Health Consequences or Impact	A05	Mechanical Problem		
									A0504	Leak/Splash		
					CT2985	Active	E2403	No Clinical Signs, Symptoms or Conditions	A05	Mechanical Problem		
									A0504	Leak/Splash		
							F26	No Health Consequences or Impact	A05	Mechanical Problem		
									A0504	Leak/Splash		
					CT3001	Implantable metal	E2403	No Clinical Signs, Symptoms or Conditions	A05	Mechanical Problem		
									A0504	Leak/Splash		
							F26	No Health Consequences or Impact	A05	Mechanical Problem		
									A0504	Leak/Splash		
					CT3002	Implantable synthetic polymer	E2403	No Clinical Signs, Symptoms or Conditions	A05	Mechanical Problem		
									A0504	Leak/Splash		
							F26	No Health Consequences or Impact	A05	Mechanical Problem		
									A0504	Leak/Splash		
					CT3006	Implantable inorganic material	E2403	No Clinical Signs, Symptoms or Conditions	A05	Mechanical Problem		
									A0504	Leak/Splash		
							F26	No Health Consequences or Impact	A05	Mechanical Problem		
									A0504	Leak/Splash		
					CT314	Electrically powered	E2403	No Clinical Signs, Symptoms or Conditions	A05	Mechanical Problem		
									A0504	Leak/Splash		
							F26	No Health Consequences or Impact	A05	Mechanical Problem		
									A0504	Leak/Splash		
					CT321	Long-term surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A05	Mechanical Problem		
									A0504	Leak/Splash		
							F26	No Health Consequences or Impact	A05	Mechanical Problem		
									A0504	Leak/Splash		
					CT981	Single use	E2403	No Clinical Signs, Symptoms or Conditions	A05	Mechanical Problem		
									A0504	Leak/Splash		
							F26	No Health Consequences or Impact	A05	Mechanical Problem		
									A0504	Leak/Splash		
					CT983	Surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A05	Mechanical Problem		
									A0504	Leak/Splash		
							F26	No Health Consequences or Impact	A05	Mechanical Problem		
									A0504	Leak/Splash		
13176	2023	All other reportable incidents	35223	Endocardial/interventricular septal pacing lead	CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070102	High Capture Threshold		
									A070901	Decreased Sensitivity		
									A070909	Oversensing		
									A090801	Signal Artifact/Noise		
							F12	Serious Injury/ Illness/ Impairment	A070102	High Capture Threshold		
									A070901	Decreased Sensitivity		
									A070909	Oversensing		
									A090801	Signal Artifact/Noise		
							F23	Unexpected Medical Intervention	A070102	High Capture Threshold		
									A070901	Decreased Sensitivity		
									A070909	Oversensing		
									A090801	Signal Artifact/Noise		
13177	2023		47265	Dual-chamber implantable pacemaker, rate-responsive	CT141	Electrodes/Leads and associated devices			A0705	Battery Problem		
					CT145	Cardiovascular devices			A0705	Battery Problem		
					CT1684	Cardiac pacemakers and associated devices			A0705	Battery Problem		
					CT1685	Cardiac pacemakers			A0705	Battery Problem		
					CT233	Surgical			A0705	Battery Problem		
					CT2406	Implantable			A0705	Battery Problem		
					CT2407	Active implantable			A0705	Battery Problem		
					CT257	Implantable cardiac pacemakers			A0705	Battery Problem		
					CT2727	Active implant devices/ancillaries			A0705	Battery Problem		
					CT2750	Implantable cardiac electrical stimulators and associated devices			A0705	Battery Problem		
					CT2866	Implantable cardiac electrical stimulators			A0705	Battery Problem		

					CT2985	Active			A0705	Battery Problem		
					CT3001	Implantable metal			A0705	Battery Problem		
					CT3002	Implantable synthetic polymer			A0705	Battery Problem		
					CT3006	Implantable inorganic material			A0705	Battery Problem		
					CT314	Electrically powered			A0705	Battery Problem		
					CT321	Long-term surgical invasive			A0705	Battery Problem		
					CT449	Generators and associated devices			A0705	Battery Problem		
					CT981	Single use			A0705	Battery Problem		
					CT983	Surgical invasive			A0705	Battery Problem		
13178	2023	All other reportable incidents	36241	Pacemaker/defibrillator lead	CT1684	Cardiac pacemakers and associated devices	E2336	Shock	A070909	Oversensing		
									A090801	Signal Artifact/Noise		
							F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A070909	Oversensing		
									A090801	Signal Artifact/Noise		
13179	2023	All other reportable incidents	47270	Cardiac resynchronization therapy implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E2336	Shock	A070909	Oversensing		
									A0712	Pacing Problem		
									A071302	Inappropriate/Inadequate Shock/Stimulation		
									A090801	Signal Artifact/Noise		
							F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A070909	Oversensing		
									A0712	Pacing Problem		
									A071302	Inappropriate/Inadequate Shock/Stimulation		
									A090801	Signal Artifact/Noise		
13180	2023	All other reportable incidents	35223	Endocardial/interventricular septal pacing lead	CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A0701	Capturing Problem		
									A072201	High Impedance		
							F1901	Additional Surgery	A0701	Capturing Problem		
									A072201	High Impedance		
13181	2023	All other reportable incidents	47265	Dual-chamber implantable pacemaker, rate-responsive	CT141	Electrodes/Leads and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	16/05/2023	16/05/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	16/05/2023	16/05/2023
					CT145	Cardiovascular devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	16/05/2023	16/05/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	16/05/2023	16/05/2023
					CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	16/05/2023	16/05/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	16/05/2023	16/05/2023
					CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	16/05/2023	16/05/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	16/05/2023	16/05/2023
					CT233	Surgical	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	16/05/2023	16/05/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	16/05/2023	16/05/2023
					CT2406	Implantable	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	16/05/2023	16/05/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	16/05/2023	16/05/2023
					CT2407	Active implantable	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	16/05/2023	16/05/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	16/05/2023	16/05/2023
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	16/05/2023	16/05/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	16/05/2023	16/05/2023
					CT2727	Active implant devices/ancillaries	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	16/05/2023	16/05/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	16/05/2023	16/05/2023
					CT2750	Implantable cardiac electrical stimulators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	16/05/2023	16/05/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	16/05/2023	16/05/2023
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	16/05/2023	16/05/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	16/05/2023	16/05/2023
					CT2985	Active	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	16/05/2023	16/05/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	16/05/2023	16/05/2023
					CT3001	Implantable metal	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	16/05/2023	16/05/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	16/05/2023	16/05/2023
					CT3002	Implantable synthetic polymer	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	16/05/2023	16/05/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	16/05/2023	16/05/2023
					CT3006	Implantable inorganic material	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	16/05/2023	16/05/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	16/05/2023	16/05/2023
					CT314	Electrically powered	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	16/05/2023	16/05/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	16/05/2023	16/05/2023
					CT321	Long-term surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	16/05/2023	16/05/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	16/05/2023	16/05/2023
					CT449	Generators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	16/05/2023	16/05/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	16/05/2023	16/05/2023
					CT981	Single use	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	16/05/2023	16/05/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	16/05/2023	16/05/2023
					CT983	Surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	16/05/2023	16/05/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	16/05/2023	16/05/2023
13182	2023	All other reportable incidents	35223	Endocardial/interventricular septal pacing lead	CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A072201	High Impedance		
							F1904	Device Repositioning	A072201	High Impedance		
							F1905	Device Revision or Replacement	A072201	High Impedance		
13183	2023	All other reportable incidents	35852	Single-chamber implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	28/02/2023	28/02/2023
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	28/02/2023	28/02/2023
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	28/02/2023	28/02/2023
13184	2023	All other reportable incidents	35039	Epicardial pacing lead	CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070102	High Capture Threshold		
							F08	Hospitalization or Prolonged Hospitalization	A070102	High Capture Threshold		
							F12	Serious Injury/ Illness/ Impairment	A070102	High Capture Threshold		
							F1905	Device Revision or Replacement	A070102	High Capture Threshold		
							F23	Unexpected Medical Intervention	A070102	High Capture Threshold		
13185	2023		37265	Dual-chamber implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E2311	Discomfort	A070909	Oversensing		
									A071302	Inappropriate/Inadequate Shock/Stimulation		
							F12	Serious Injury/ Illness/ Impairment	A070909	Oversensing		
									A071302	Inappropriate/Inadequate Shock/Stimulation		
13186	2023	All other reportable incidents	47265	Dual-chamber implantable pacemaker, rate-responsive	CT1685	Cardiac pacemakers	E0112	Dizziness	A071204	No Pacing	24/02/2023	24/02/2023
									A071205	Pacemaker Found in Back-Up Mode	24/02/2023	24/02/2023
							F1901	Additional Surgery	A071204	No Pacing	24/02/2023	24/02/2023
									A071205	Pacemaker Found in Back-Up Mode	24/02/2023	24/02/2023
					CT257	Implantable cardiac pacemakers	E0112	Dizziness	A071204	No Pacing	24/02/2023	24/02/2023
									A071205	Pacemaker Found in Back-Up Mode	24/02/2023	24/02/2023
							F1901	Additional Surgery	A071204	No Pacing	24/02/2023	24/02/2023
									A071205	Pacemaker Found in Back-Up Mode	24/02/2023	24/02/2023

					CT2866	Implantable cardiac electrical stimulators	E0112	Dizziness	A071204	No Pacing	24/02/2023	24/02/2023
									A071205	Pacemaker Found in Back-Up Mode	24/02/2023	24/02/2023
							F1901	Additional Surgery	A071204	No Pacing	24/02/2023	24/02/2023
									A071205	Pacemaker Found in Back-Up Mode	24/02/2023	24/02/2023
13187	2023	All other reportable incidents	47265	Dual-chamber implantable pacemaker, rate-responsive	CT141	Electrodes/Leads and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A160601	Premature Elective Replacement Indicator		
							F1903	Device Explantation	A160601	Premature Elective Replacement Indicator		
					CT145	Cardiovascular devices	E2403	No Clinical Signs, Symptoms or Conditions	A160601	Premature Elective Replacement Indicator		
							F1903	Device Explantation	A160601	Premature Elective Replacement Indicator		
					CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A160601	Premature Elective Replacement Indicator		
							F1903	Device Explantation	A160601	Premature Elective Replacement Indicator		
					CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A160601	Premature Elective Replacement Indicator		
							F1903	Device Explantation	A160601	Premature Elective Replacement Indicator		
					CT233	Surgical	E2403	No Clinical Signs, Symptoms or Conditions	A160601	Premature Elective Replacement Indicator		
							F1903	Device Explantation	A160601	Premature Elective Replacement Indicator		
					CT2406	Implantable	E2403	No Clinical Signs, Symptoms or Conditions	A160601	Premature Elective Replacement Indicator		
							F1903	Device Explantation	A160601	Premature Elective Replacement Indicator		
					CT2407	Active implantable	E2403	No Clinical Signs, Symptoms or Conditions	A160601	Premature Elective Replacement Indicator		
							F1903	Device Explantation	A160601	Premature Elective Replacement Indicator		
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A160601	Premature Elective Replacement Indicator		
							F1903	Device Explantation	A160601	Premature Elective Replacement Indicator		
					CT2727	Active implant devices/ancillaries	E2403	No Clinical Signs, Symptoms or Conditions	A160601	Premature Elective Replacement Indicator		
							F1903	Device Explantation	A160601	Premature Elective Replacement Indicator		
					CT2750	Implantable cardiac electrical stimulators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A160601	Premature Elective Replacement Indicator		
							F1903	Device Explantation	A160601	Premature Elective Replacement Indicator		
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A160601	Premature Elective Replacement Indicator		
							F1903	Device Explantation	A160601	Premature Elective Replacement Indicator		
					CT2985	Active	E2403	No Clinical Signs, Symptoms or Conditions	A160601	Premature Elective Replacement Indicator		
							F1903	Device Explantation	A160601	Premature Elective Replacement Indicator		
					CT3001	Implantable metal	E2403	No Clinical Signs, Symptoms or Conditions	A160601	Premature Elective Replacement Indicator		
							F1903	Device Explantation	A160601	Premature Elective Replacement Indicator		
					CT3002	Implantable synthetic polymer	E2403	No Clinical Signs, Symptoms or Conditions	A160601	Premature Elective Replacement Indicator		
							F1903	Device Explantation	A160601	Premature Elective Replacement Indicator		
					CT3006	Implantable inorganic material	E2403	No Clinical Signs, Symptoms or Conditions	A160601	Premature Elective Replacement Indicator		
							F1903	Device Explantation	A160601	Premature Elective Replacement Indicator		
					CT314	Electrically powered	E2403	No Clinical Signs, Symptoms or Conditions	A160601	Premature Elective Replacement Indicator		
							F1903	Device Explantation	A160601	Premature Elective Replacement Indicator		
					CT321	Long-term surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A160601	Premature Elective Replacement Indicator		
							F1903	Device Explantation	A160601	Premature Elective Replacement Indicator		
					CT449	Generators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A160601	Premature Elective Replacement Indicator		
							F1903	Device Explantation	A160601	Premature Elective Replacement Indicator		
					CT981	Single use	E2403	No Clinical Signs, Symptoms or Conditions	A160601	Premature Elective Replacement Indicator		
							F1903	Device Explantation	A160601	Premature Elective Replacement Indicator		
					CT983	Surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A160601	Premature Elective Replacement Indicator		
							F1903	Device Explantation	A160601	Premature Elective Replacement Indicator		
13188	2023	All other reportable incidents	47265	Dual-chamber implantable pacemaker, rate-responsive	CT141	Electrodes/Leads and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT145	Cardiovascular devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT233	Surgical	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT2406	Implantable	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT2407	Active implantable	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT2727	Active implant devices/ancillaries	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT2750	Implantable cardiac electrical stimulators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT2985	Active	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT3001	Implantable metal	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT3002	Implantable synthetic polymer	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT3006	Implantable inorganic material	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT314	Electrically powered	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT321	Long-term surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT449	Generators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT981	Single use	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT983	Surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
13189	2023	All other reportable incidents	35852	Single-chamber implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	17/05/2023	17/05/2023
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	17/05/2023	17/05/2023
							F12	Serious Injury/ Illness/ Impairment	A070504	Premature Discharge of Battery	17/05/2023	17/05/2023
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	17/05/2023	17/05/2023

13190	2023	All other reportable incidents	47265	Dual-chamber implantable pacemaker, rate-responsive	CT141	Electrodes/Leads and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A0908	Incorrect, Inadequate or Imprecise Result or Readings	12/05/2023	12/05/2023
							F08	Hospitalization or Prolonged Hospitalization	A0908	Incorrect, Inadequate or Imprecise Result or Readings	12/05/2023	12/05/2023
							F1905	Device Revision or Replacement	A0908	Incorrect, Inadequate or Imprecise Result or Readings	12/05/2023	12/05/2023
					CT145	Cardiovascular devices	E2403	No Clinical Signs, Symptoms or Conditions	A0908	Incorrect, Inadequate or Imprecise Result or Readings	12/05/2023	12/05/2023
							F08	Hospitalization or Prolonged Hospitalization	A0908	Incorrect, Inadequate or Imprecise Result or Readings	12/05/2023	12/05/2023
							F1905	Device Revision or Replacement	A0908	Incorrect, Inadequate or Imprecise Result or Readings	12/05/2023	12/05/2023
					CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A0908	Incorrect, Inadequate or Imprecise Result or Readings	12/05/2023	12/05/2023
							F08	Hospitalization or Prolonged Hospitalization	A0908	Incorrect, Inadequate or Imprecise Result or Readings	12/05/2023	12/05/2023
							F1905	Device Revision or Replacement	A0908	Incorrect, Inadequate or Imprecise Result or Readings	12/05/2023	12/05/2023
					CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A0908	Incorrect, Inadequate or Imprecise Result or Readings	12/05/2023	12/05/2023
							F08	Hospitalization or Prolonged Hospitalization	A0908	Incorrect, Inadequate or Imprecise Result or Readings	12/05/2023	12/05/2023
							F1905	Device Revision or Replacement	A0908	Incorrect, Inadequate or Imprecise Result or Readings	12/05/2023	12/05/2023
					CT233	Surgical	E2403	No Clinical Signs, Symptoms or Conditions	A0908	Incorrect, Inadequate or Imprecise Result or Readings	12/05/2023	12/05/2023
							F08	Hospitalization or Prolonged Hospitalization	A0908	Incorrect, Inadequate or Imprecise Result or Readings	12/05/2023	12/05/2023
							F1905	Device Revision or Replacement	A0908	Incorrect, Inadequate or Imprecise Result or Readings	12/05/2023	12/05/2023
					CT2406	Implantable	E2403	No Clinical Signs, Symptoms or Conditions	A0908	Incorrect, Inadequate or Imprecise Result or Readings	12/05/2023	12/05/2023
							F08	Hospitalization or Prolonged Hospitalization	A0908	Incorrect, Inadequate or Imprecise Result or Readings	12/05/2023	12/05/2023
							F1905	Device Revision or Replacement	A0908	Incorrect, Inadequate or Imprecise Result or Readings	12/05/2023	12/05/2023
					CT2407	Active implantable	E2403	No Clinical Signs, Symptoms or Conditions	A0908	Incorrect, Inadequate or Imprecise Result or Readings	12/05/2023	12/05/2023
							F08	Hospitalization or Prolonged Hospitalization	A0908	Incorrect, Inadequate or Imprecise Result or Readings	12/05/2023	12/05/2023
							F1905	Device Revision or Replacement	A0908	Incorrect, Inadequate or Imprecise Result or Readings	12/05/2023	12/05/2023
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A0908	Incorrect, Inadequate or Imprecise Result or Readings	12/05/2023	12/05/2023
							F08	Hospitalization or Prolonged Hospitalization	A0908	Incorrect, Inadequate or Imprecise Result or Readings	12/05/2023	12/05/2023
							F1905	Device Revision or Replacement	A0908	Incorrect, Inadequate or Imprecise Result or Readings	12/05/2023	12/05/2023
					CT2727	Active implant devices/ancillaries	E2403	No Clinical Signs, Symptoms or Conditions	A0908	Incorrect, Inadequate or Imprecise Result or Readings	12/05/2023	12/05/2023
							F08	Hospitalization or Prolonged Hospitalization	A0908	Incorrect, Inadequate or Imprecise Result or Readings	12/05/2023	12/05/2023
							F1905	Device Revision or Replacement	A0908	Incorrect, Inadequate or Imprecise Result or Readings	12/05/2023	12/05/2023
					CT2750	Implantable cardiac electrical stimulators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A0908	Incorrect, Inadequate or Imprecise Result or Readings	12/05/2023	12/05/2023
							F08	Hospitalization or Prolonged Hospitalization	A0908	Incorrect, Inadequate or Imprecise Result or Readings	12/05/2023	12/05/2023
							F1905	Device Revision or Replacement	A0908	Incorrect, Inadequate or Imprecise Result or Readings	12/05/2023	12/05/2023
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A0908	Incorrect, Inadequate or Imprecise Result or Readings	12/05/2023	12/05/2023
							F08	Hospitalization or Prolonged Hospitalization	A0908	Incorrect, Inadequate or Imprecise Result or Readings	12/05/2023	12/05/2023
							F1905	Device Revision or Replacement	A0908	Incorrect, Inadequate or Imprecise Result or Readings	12/05/2023	12/05/2023
					CT2985	Active	E2403	No Clinical Signs, Symptoms or Conditions	A0908	Incorrect, Inadequate or Imprecise Result or Readings	12/05/2023	12/05/2023
							F08	Hospitalization or Prolonged Hospitalization	A0908	Incorrect, Inadequate or Imprecise Result or Readings	12/05/2023	12/05/2023
							F1905	Device Revision or Replacement	A0908	Incorrect, Inadequate or Imprecise Result or Readings	12/05/2023	12/05/2023
					CT3001	Implantable metal	E2403	No Clinical Signs, Symptoms or Conditions	A0908	Incorrect, Inadequate or Imprecise Result or Readings	12/05/2023	12/05/2023
							F08	Hospitalization or Prolonged Hospitalization	A0908	Incorrect, Inadequate or Imprecise Result or Readings	12/05/2023	12/05/2023
							F1905	Device Revision or Replacement	A0908	Incorrect, Inadequate or Imprecise Result or Readings	12/05/2023	12/05/2023
					CT3002	Implantable synthetic polymer	E2403	No Clinical Signs, Symptoms or Conditions	A0908	Incorrect, Inadequate or Imprecise Result or Readings	12/05/2023	12/05/2023
							F08	Hospitalization or Prolonged Hospitalization	A0908	Incorrect, Inadequate or Imprecise Result or Readings	12/05/2023	12/05/2023
							F1905	Device Revision or Replacement	A0908	Incorrect, Inadequate or Imprecise Result or Readings	12/05/2023	12/05/2023
					CT3006	Implantable inorganic material	E2403	No Clinical Signs, Symptoms or Conditions	A0908	Incorrect, Inadequate or Imprecise Result or Readings	12/05/2023	12/05/2023
							F08	Hospitalization or Prolonged Hospitalization	A0908	Incorrect, Inadequate or Imprecise Result or Readings	12/05/2023	12/05/2023
							F1905	Device Revision or Replacement	A0908	Incorrect, Inadequate or Imprecise Result or Readings	12/05/2023	12/05/2023
					CT314	Electrically powered	E2403	No Clinical Signs, Symptoms or Conditions	A0908	Incorrect, Inadequate or Imprecise Result or Readings	12/05/2023	12/05/2023
							F08	Hospitalization or Prolonged Hospitalization	A0908	Incorrect, Inadequate or Imprecise Result or Readings	12/05/2023	12/05/2023
							F1905	Device Revision or Replacement	A0908	Incorrect, Inadequate or Imprecise Result or Readings	12/05/2023	12/05/2023
					CT321	Long-term surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A0908	Incorrect, Inadequate or Imprecise Result or Readings	12/05/2023	12/05/2023
							F08	Hospitalization or Prolonged Hospitalization	A0908	Incorrect, Inadequate or Imprecise Result or Readings	12/05/2023	12/05/2023

							F1905	Device Revision or Replacement	A0908	Incorrect, Inadequate or Imprecise Result or Readings	12/05/2023	12/05/2023
					CT449	Generators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A0908	Incorrect, Inadequate or Imprecise Result or Readings	12/05/2023	12/05/2023
							F08	Hospitalization or Prolonged Hospitalization	A0908	Incorrect, Inadequate or Imprecise Result or Readings	12/05/2023	12/05/2023
							F1905	Device Revision or Replacement	A0908	Incorrect, Inadequate or Imprecise Result or Readings	12/05/2023	12/05/2023
					CT981	Single use	E2403	No Clinical Signs, Symptoms or Conditions	A0908	Incorrect, Inadequate or Imprecise Result or Readings	12/05/2023	12/05/2023
							F08	Hospitalization or Prolonged Hospitalization	A0908	Incorrect, Inadequate or Imprecise Result or Readings	12/05/2023	12/05/2023
							F1905	Device Revision or Replacement	A0908	Incorrect, Inadequate or Imprecise Result or Readings	12/05/2023	12/05/2023
					CT983	Surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A0908	Incorrect, Inadequate or Imprecise Result or Readings	12/05/2023	12/05/2023
							F08	Hospitalization or Prolonged Hospitalization	A0908	Incorrect, Inadequate or Imprecise Result or Readings	12/05/2023	12/05/2023
							F1905	Device Revision or Replacement	A0908	Incorrect, Inadequate or Imprecise Result or Readings	12/05/2023	12/05/2023
13191	2023	All other reportable incidents	47265	Dual-chamber implantable pacemaker, rate-responsive	CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
13192	2023	All other reportable incidents	47265	Dual-chamber implantable pacemaker, rate-responsive			E2403	No Clinical Signs, Symptoms or Conditions	A0908	Incorrect, Inadequate or Imprecise Result or Readings	18/05/2023	18/05/2023
							F08	Hospitalization or Prolonged Hospitalization	A0908	Incorrect, Inadequate or Imprecise Result or Readings	18/05/2023	18/05/2023
							F1905	Device Revision or Replacement	A0908	Incorrect, Inadequate or Imprecise Result or Readings	18/05/2023	18/05/2023
13193	2023	All other reportable incidents	47267	Single-chamber implantable pacemaker, rate-responsive	CT141	Electrodes/Leads and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
							F24	Insufficient Health Impact Information	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
					CT145	Cardiovascular devices	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
							F24	Insufficient Health Impact Information	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
					CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
							F24	Insufficient Health Impact Information	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
					CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
							F24	Insufficient Health Impact Information	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
					CT233	Surgical	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
							F24	Insufficient Health Impact Information	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
					CT2406	Implantable	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
							F24	Insufficient Health Impact Information	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
					CT2407	Active implantable	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
							F24	Insufficient Health Impact Information	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
							F24	Insufficient Health Impact Information	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
					CT2727	Active implant devices/ancillaries	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
							F24	Insufficient Health Impact Information	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
					CT2750	Implantable cardiac electrical stimulators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
							F24	Insufficient Health Impact Information	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
							F24	Insufficient Health Impact Information	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
					CT2985	Active	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
							F24	Insufficient Health Impact Information	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
					CT3001	Implantable metal	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
							F24	Insufficient Health Impact Information	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
					CT3002	Implantable synthetic polymer	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
							F24	Insufficient Health Impact Information	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
					CT3006	Implantable inorganic material	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
							F24	Insufficient Health Impact Information	A0705	Battery Problem		

								A070504	Premature Discharge of Battery			
				CT314	Electrically powered	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem			
								A070504	Premature Discharge of Battery			
						F24	Insufficient Health Impact Information	A0705	Battery Problem			
								A070504	Premature Discharge of Battery			
				CT321	Long-term surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem			
								A070504	Premature Discharge of Battery			
						F24	Insufficient Health Impact Information	A0705	Battery Problem			
								A070504	Premature Discharge of Battery			
				CT449	Generators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem			
								A070504	Premature Discharge of Battery			
						F24	Insufficient Health Impact Information	A0705	Battery Problem			
								A070504	Premature Discharge of Battery			
				CT981	Single use	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem			
								A070504	Premature Discharge of Battery			
						F24	Insufficient Health Impact Information	A0705	Battery Problem			
								A070504	Premature Discharge of Battery			
				CT983	Surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem			
								A070504	Premature Discharge of Battery			
						F24	Insufficient Health Impact Information	A0705	Battery Problem			
								A070504	Premature Discharge of Battery			
13194	2023	All other reportable incidents	47205	Cardiac pulse generator programmer	CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A0203	Defective Device		
									A150105	Key or Button Unresponsive/not Working		
						F26	No Health Consequences or Impact	A0203	Defective Device			
								A150105	Key or Button Unresponsive/not Working			
13195	2023	All other reportable incidents	47265	Dual-chamber implantable pacemaker, rate-responsive	CT141	Electrodes/Leads and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	02/05/2023	02/05/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	02/05/2023	02/05/2023
					CT145	Cardiovascular devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	02/05/2023	02/05/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	02/05/2023	02/05/2023
					CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	02/05/2023	02/05/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	02/05/2023	02/05/2023
					CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	02/05/2023	02/05/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	02/05/2023	02/05/2023
					CT233	Surgical	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	02/05/2023	02/05/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	02/05/2023	02/05/2023
					CT2406	Implantable	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	02/05/2023	02/05/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	02/05/2023	02/05/2023
					CT2407	Active implantable	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	02/05/2023	02/05/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	02/05/2023	02/05/2023
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	02/05/2023	02/05/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	02/05/2023	02/05/2023
					CT2727	Active implant devices/ancillaries	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	02/05/2023	02/05/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	02/05/2023	02/05/2023
					CT2750	Implantable cardiac electrical stimulators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	02/05/2023	02/05/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	02/05/2023	02/05/2023
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	02/05/2023	02/05/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	02/05/2023	02/05/2023
					CT2985	Active	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	02/05/2023	02/05/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	02/05/2023	02/05/2023
					CT3001	Implantable metal	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	02/05/2023	02/05/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	02/05/2023	02/05/2023
					CT3002	Implantable synthetic polymer	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	02/05/2023	02/05/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	02/05/2023	02/05/2023
					CT3006	Implantable inorganic material	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	02/05/2023	02/05/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	02/05/2023	02/05/2023
					CT314	Electrically powered	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	02/05/2023	02/05/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	02/05/2023	02/05/2023
					CT321	Long-term surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	02/05/2023	02/05/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	02/05/2023	02/05/2023
					CT449	Generators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	02/05/2023	02/05/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	02/05/2023	02/05/2023
					CT981	Single use	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	02/05/2023	02/05/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	02/05/2023	02/05/2023
					CT983	Surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	02/05/2023	02/05/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	02/05/2023	02/05/2023
13196	2023	Death	37265	Dual-chamber implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E0611	Heart Failure/Congestive Heart Failure	A24	No Apparent Device Problem		
							F02	Death	A24	No Apparent Device Problem		
13197	2023	All other reportable incidents	47205	Cardiac pulse generator programmer	CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A150105	Key or Button Unresponsive/not Working		
							F26	No Health Consequences or Impact	A150105	Key or Button Unresponsive/not Working		
13198	2023	All other reportable incidents	47267	Single-chamber implantable pacemaker, rate-responsive	CT141	Electrodes/Leads and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	01/02/2023	01/02/2023
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	01/02/2023	01/02/2023
							F1903	Device Explantation	A070504	Premature Discharge of Battery	01/02/2023	01/02/2023
					CT145	Cardiovascular devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	01/02/2023	01/02/2023
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	01/02/2023	01/02/2023
							F1903	Device Explantation	A070504	Premature Discharge of Battery	01/02/2023	01/02/2023
					CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	01/02/2023	01/02/2023
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	01/02/2023	01/02/2023
							F1903	Device Explantation	A070504	Premature Discharge of Battery	01/02/2023	01/02/2023
					CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	01/02/2023	01/02/2023
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	01/02/2023	01/02/2023
							F1903	Device Explantation	A070504	Premature Discharge of Battery	01/02/2023	01/02/2023
					CT233	Surgical	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	01/02/2023	01/02/2023
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	01/02/2023	01/02/2023
							F1903	Device Explantation	A070504	Premature Discharge of Battery	01/02/2023	01/02/2023
					CT2406	Implantable	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	01/02/2023	01/02/2023
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	01/02/2023	01/02/2023
							F1903	Device Explantation	A070504	Premature Discharge of Battery	01/02/2023	01/02/2023
					CT2407	Active implantable	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	01/02/2023	01/02/2023

							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	01/02/2023	01/02/2023
							F1903	Device Explantation	A070504	Premature Discharge of Battery	01/02/2023	01/02/2023
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	01/02/2023	01/02/2023
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	01/02/2023	01/02/2023
							F1903	Device Explantation	A070504	Premature Discharge of Battery	01/02/2023	01/02/2023
					CT2727	Active implant devices/ancillaries	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	01/02/2023	01/02/2023
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	01/02/2023	01/02/2023
							F1903	Device Explantation	A070504	Premature Discharge of Battery	01/02/2023	01/02/2023
					CT2750	Implantable cardiac electrical stimulators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	01/02/2023	01/02/2023
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	01/02/2023	01/02/2023
							F1903	Device Explantation	A070504	Premature Discharge of Battery	01/02/2023	01/02/2023
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	01/02/2023	01/02/2023
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	01/02/2023	01/02/2023
							F1903	Device Explantation	A070504	Premature Discharge of Battery	01/02/2023	01/02/2023
					CT2985	Active	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	01/02/2023	01/02/2023
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	01/02/2023	01/02/2023
							F1903	Device Explantation	A070504	Premature Discharge of Battery	01/02/2023	01/02/2023
					CT3001	Implantable metal	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	01/02/2023	01/02/2023
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	01/02/2023	01/02/2023
							F1903	Device Explantation	A070504	Premature Discharge of Battery	01/02/2023	01/02/2023
					CT3002	Implantable synthetic polymer	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	01/02/2023	01/02/2023
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	01/02/2023	01/02/2023
							F1903	Device Explantation	A070504	Premature Discharge of Battery	01/02/2023	01/02/2023
					CT3006	Implantable inorganic material	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	01/02/2023	01/02/2023
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	01/02/2023	01/02/2023
							F1903	Device Explantation	A070504	Premature Discharge of Battery	01/02/2023	01/02/2023
					CT314	Electrically powered	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	01/02/2023	01/02/2023
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	01/02/2023	01/02/2023
							F1903	Device Explantation	A070504	Premature Discharge of Battery	01/02/2023	01/02/2023
					CT321	Long-term surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	01/02/2023	01/02/2023
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	01/02/2023	01/02/2023
							F1903	Device Explantation	A070504	Premature Discharge of Battery	01/02/2023	01/02/2023
					CT449	Generators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	01/02/2023	01/02/2023
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	01/02/2023	01/02/2023
							F1903	Device Explantation	A070504	Premature Discharge of Battery	01/02/2023	01/02/2023
					CT981	Single use	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	01/02/2023	01/02/2023
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	01/02/2023	01/02/2023
							F1903	Device Explantation	A070504	Premature Discharge of Battery	01/02/2023	01/02/2023
					CT983	Surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	01/02/2023	01/02/2023
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	01/02/2023	01/02/2023
							F1903	Device Explantation	A070504	Premature Discharge of Battery	01/02/2023	01/02/2023
13199	2023		47265	Dual-chamber implantable pacemaker, rate-responsive	CT141	Electrodes/Leads and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	15/03/2023	15/03/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	15/03/2023	15/03/2023
					CT145	Cardiovascular devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	15/03/2023	15/03/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	15/03/2023	15/03/2023
					CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	15/03/2023	15/03/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	15/03/2023	15/03/2023
					CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	15/03/2023	15/03/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	15/03/2023	15/03/2023
					CT233	Surgical	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	15/03/2023	15/03/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	15/03/2023	15/03/2023
					CT2406	Implantable	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	15/03/2023	15/03/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	15/03/2023	15/03/2023
					CT2407	Active implantable	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	15/03/2023	15/03/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	15/03/2023	15/03/2023
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	15/03/2023	15/03/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	15/03/2023	15/03/2023
					CT2727	Active implant devices/ancillaries	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	15/03/2023	15/03/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	15/03/2023	15/03/2023
					CT2750	Implantable cardiac electrical stimulators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	15/03/2023	15/03/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	15/03/2023	15/03/2023
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	15/03/2023	15/03/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	15/03/2023	15/03/2023
					CT2985	Active	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	15/03/2023	15/03/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	15/03/2023	15/03/2023
					CT3001	Implantable metal	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	15/03/2023	15/03/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	15/03/2023	15/03/2023
					CT3002	Implantable synthetic polymer	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	15/03/2023	15/03/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	15/03/2023	15/03/2023
					CT3006	Implantable inorganic material	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	15/03/2023	15/03/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	15/03/2023	15/03/2023
					CT314	Electrically powered	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	15/03/2023	15/03/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	15/03/2023	15/03/2023
					CT321	Long-term surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	15/03/2023	15/03/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	15/03/2023	15/03/2023
					CT449	Generators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	15/03/2023	15/03/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	15/03/2023	15/03/2023
					CT981	Single use	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	15/03/2023	15/03/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	15/03/2023	15/03/2023
					CT983	Surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	15/03/2023	15/03/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	15/03/2023	15/03/2023
13200	2023		47265	Dual-chamber implantable pacemaker, rate-responsive	CT141	Electrodes/Leads and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	22/03/2023	22/03/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	22/03/2023	22/03/2023
					CT145	Cardiovascular devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	22/03/2023	22/03/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	22/03/2023	22/03/2023
					CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	22/03/2023	22/03/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	22/03/2023	22/03/2023
					CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	22/03/2023	22/03/2023

							F1901	Additional Surgery	A24	No Apparent Device Problem	22/03/2023	22/03/2023
					CT233	Surgical	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	22/03/2023	22/03/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	22/03/2023	22/03/2023
					CT2406	Implantable	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	22/03/2023	22/03/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	22/03/2023	22/03/2023
					CT2407	Active implantable	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	22/03/2023	22/03/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	22/03/2023	22/03/2023
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	22/03/2023	22/03/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	22/03/2023	22/03/2023
					CT2727	Active implant devices/ancillaries	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	22/03/2023	22/03/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	22/03/2023	22/03/2023
					CT2750	Implantable cardiac electrical stimulators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	22/03/2023	22/03/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	22/03/2023	22/03/2023
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	22/03/2023	22/03/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	22/03/2023	22/03/2023
					CT2985	Active	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	22/03/2023	22/03/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	22/03/2023	22/03/2023
					CT3001	Implantable metal	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	22/03/2023	22/03/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	22/03/2023	22/03/2023
					CT3002	Implantable synthetic polymer	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	22/03/2023	22/03/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	22/03/2023	22/03/2023
					CT3006	Implantable inorganic material	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	22/03/2023	22/03/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	22/03/2023	22/03/2023
					CT314	Electrically powered	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	22/03/2023	22/03/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	22/03/2023	22/03/2023
					CT321	Long-term surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	22/03/2023	22/03/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	22/03/2023	22/03/2023
					CT449	Generators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	22/03/2023	22/03/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	22/03/2023	22/03/2023
					CT981	Single use	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	22/03/2023	22/03/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	22/03/2023	22/03/2023
					CT983	Surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	22/03/2023	22/03/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	22/03/2023	22/03/2023
13201	2023	All other reportable incidents	36241	Pacemaker/defibrillator lead	CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A072201	High Impedance		
							F26	No Health Consequences or Impact	A072201	High Impedance		
13202	2023	All other reportable incidents	17882	Physiologic-monitoring defibrillation system	CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A07	Electrical /Electronic Property Problem		
							F27	Problem Identified During Non-Clinical Procedure	A07	Electrical /Electronic Property Problem		
13203	2023		47265	Dual-chamber implantable pacemaker, rate-responsive	CT141	Electrodes/Leads and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem		
							F1901	Additional Surgery	A24	No Apparent Device Problem		
					CT145	Cardiovascular devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem		
							F1901	Additional Surgery	A24	No Apparent Device Problem		
					CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem		
							F1901	Additional Surgery	A24	No Apparent Device Problem		
					CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem		
							F1901	Additional Surgery	A24	No Apparent Device Problem		
					CT233	Surgical	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem		
							F1901	Additional Surgery	A24	No Apparent Device Problem		
					CT2406	Implantable	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem		
							F1901	Additional Surgery	A24	No Apparent Device Problem		
					CT2407	Active implantable	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem		
							F1901	Additional Surgery	A24	No Apparent Device Problem		
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem		
							F1901	Additional Surgery	A24	No Apparent Device Problem		
					CT2727	Active implant devices/ancillaries	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem		
							F1901	Additional Surgery	A24	No Apparent Device Problem		
					CT2750	Implantable cardiac electrical stimulators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem		
							F1901	Additional Surgery	A24	No Apparent Device Problem		
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem		
							F1901	Additional Surgery	A24	No Apparent Device Problem		
					CT2985	Active	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem		
							F1901	Additional Surgery	A24	No Apparent Device Problem		
					CT3001	Implantable metal	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem		
							F1901	Additional Surgery	A24	No Apparent Device Problem		
					CT3002	Implantable synthetic polymer	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem		
							F1901	Additional Surgery	A24	No Apparent Device Problem		
					CT3006	Implantable inorganic material	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem		
							F1901	Additional Surgery	A24	No Apparent Device Problem		
					CT314	Electrically powered	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem		
							F1901	Additional Surgery	A24	No Apparent Device Problem		
					CT321	Long-term surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem		
							F1901	Additional Surgery	A24	No Apparent Device Problem		
					CT449	Generators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem		
							F1901	Additional Surgery	A24	No Apparent Device Problem		
					CT981	Single use	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem		
							F1901	Additional Surgery	A24	No Apparent Device Problem		
					CT983	Surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem		
							F1901	Additional Surgery	A24	No Apparent Device Problem		
13204	2023		47265	Dual-chamber implantable pacemaker, rate-responsive	CT141	Electrodes/Leads and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem		
							F1901	Additional Surgery	A24	No Apparent Device Problem		
					CT145	Cardiovascular devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem		
							F1901	Additional Surgery	A24	No Apparent Device Problem		
					CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem		
							F1901	Additional Surgery	A24	No Apparent Device Problem		
					CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem		
							F1901	Additional Surgery	A24	No Apparent Device Problem		
					CT233	Surgical	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem		
							F1901	Additional Surgery	A24	No Apparent Device Problem		
					CT2406	Implantable	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem		

							F1901	Additional Surgery	A24	No Apparent Device Problem		
					CT2407	Active implantable	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem		
							F1901	Additional Surgery	A24	No Apparent Device Problem		
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem		
							F1901	Additional Surgery	A24	No Apparent Device Problem		
					CT2727	Active implant devices/ancillaries	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem		
							F1901	Additional Surgery	A24	No Apparent Device Problem		
					CT2750	Implantable cardiac electrical stimulators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem		
							F1901	Additional Surgery	A24	No Apparent Device Problem		
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem		
							F1901	Additional Surgery	A24	No Apparent Device Problem		
					CT2985	Active	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem		
							F1901	Additional Surgery	A24	No Apparent Device Problem		
					CT3001	Implantable metal	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem		
							F1901	Additional Surgery	A24	No Apparent Device Problem		
					CT3002	Implantable synthetic polymer	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem		
							F1901	Additional Surgery	A24	No Apparent Device Problem		
					CT3006	Implantable inorganic material	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem		
							F1901	Additional Surgery	A24	No Apparent Device Problem		
					CT314	Electrically powered	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem		
							F1901	Additional Surgery	A24	No Apparent Device Problem		
					CT321	Long-term surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem		
							F1901	Additional Surgery	A24	No Apparent Device Problem		
					CT449	Generators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem		
							F1901	Additional Surgery	A24	No Apparent Device Problem		
					CT981	Single use	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem		
							F1901	Additional Surgery	A24	No Apparent Device Problem		
					CT983	Surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem		
							F1901	Additional Surgery	A24	No Apparent Device Problem		
13205	2023	All other reportable incidents	35223	Endocardial/interventricular septal pacing lead	CT1684	Cardiac pacemakers and associated devices	E060106	Heart Block	A070909	Oversensing		
							F12	Serious Injury/ Illness/ Impairment	A070909	Oversensing		
13206	2023		37265	Dual-chamber implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	F26	No Health Consequences or Impact	A09	Output Problem		
							F27	Problem Identified During Non-Clinical Procedure	A09	Output Problem		
13207	2023	All other reportable incidents	35223	Endocardial/interventricular septal pacing lead	CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A072201	High Impedance	23/02/2023	23/02/2023
							F08	Hospitalization or Prolonged Hospitalization	A072201	High Impedance	23/02/2023	23/02/2023
							F12	Serious Injury/ Illness/ Impairment	A072201	High Impedance	23/02/2023	23/02/2023
							F1903	Device Explantation	A072201	High Impedance	23/02/2023	23/02/2023
13208	2023	All other reportable incidents	36241	Pacemaker/defibrillator lead	CT1684	Cardiac pacemakers and associated devices	E2103	Device Overstimulation of Tissue	A051201	Device Dislodged or Dislocated		
									A070101	Failure to Capture		
									A072201	High Impedance		
							F08	Hospitalization or Prolonged Hospitalization	A051201	Device Dislodged or Dislocated		
									A070101	Failure to Capture		
									A072201	High Impedance		
							F1905	Device Revision or Replacement	A051201	Device Dislodged or Dislocated		
									A070101	Failure to Capture		
									A072201	High Impedance		
13209	2023	All other reportable incidents	47265	Dual-chamber implantable pacemaker, rate-responsive	CT141	Electrodes/Leads and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
					CT145	Cardiovascular devices	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
					CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
					CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
					CT233	Surgical	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
					CT2406	Implantable	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
					CT2407	Active implantable	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
					CT2727	Active implant devices/ancillaries	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
					CT2750	Implantable cardiac electrical stimulators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		

							F26	No Health Consequences or Impact	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
					CT2985	Active	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
					CT3001	Implantable metal	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
					CT3002	Implantable synthetic polymer	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
					CT3006	Implantable inorganic material	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
					CT314	Electrically powered	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
					CT321	Long-term surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
					CT449	Generators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
					CT981	Single use	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
					CT983	Surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
13210	2023	All other reportable incidents	35223	Endocardial/interventricular septal pacing lead	CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A07	Electrical /Electronic Property Problem		
									A072201	High Impedance		
									A090801	Signal Artifact/Noise		
							F12	Serious Injury/ Illness/ Impairment	A07	Electrical /Electronic Property Problem		
									A072201	High Impedance		
									A090801	Signal Artifact/Noise		
							F23	Unexpected Medical Intervention	A07	Electrical /Electronic Property Problem		
									A072201	High Impedance		
									A090801	Signal Artifact/Noise		
13211	2023	All other reportable incidents	37265	Dual-chamber implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	04/07/2023	04/07/2023
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	04/07/2023	04/07/2023
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	04/07/2023	04/07/2023
13212	2023	All other reportable incidents	35223	Endocardial/interventricular septal pacing lead	CT1684	Cardiac pacemakers and associated devices	E0112	Dizziness	A070101	Failure to Capture	26/05/2023	26/05/2023
									A070909	Oversensing	26/05/2023	26/05/2023
									A072201	High Impedance	26/05/2023	26/05/2023
							E011903	Syncope/Fainting	A070101	Failure to Capture	26/05/2023	26/05/2023
									A070909	Oversensing	26/05/2023	26/05/2023
									A072201	High Impedance	26/05/2023	26/05/2023
							E2002	Bruise/Contusion	A070101	Failure to Capture	26/05/2023	26/05/2023
									A070909	Oversensing	26/05/2023	26/05/2023
									A072201	High Impedance	26/05/2023	26/05/2023
							E2007	Fall	A070101	Failure to Capture	26/05/2023	26/05/2023
									A070909	Oversensing	26/05/2023	26/05/2023
									A072201	High Impedance	26/05/2023	26/05/2023
							F08	Hospitalization or Prolonged Hospitalization	A070101	Failure to Capture	26/05/2023	26/05/2023
									A070909	Oversensing	26/05/2023	26/05/2023
									A072201	High Impedance	26/05/2023	26/05/2023
							F1203	Life Threatening Illness or Injury	A070101	Failure to Capture	26/05/2023	26/05/2023
									A070909	Oversensing	26/05/2023	26/05/2023
									A072201	High Impedance	26/05/2023	26/05/2023
							F1905	Device Revision or Replacement	A070101	Failure to Capture	26/05/2023	26/05/2023
									A070909	Oversensing	26/05/2023	26/05/2023
									A072201	High Impedance	26/05/2023	26/05/2023
							F23	Unexpected Medical Intervention	A070101	Failure to Capture	26/05/2023	26/05/2023
									A070909	Oversensing	26/05/2023	26/05/2023
									A072201	High Impedance	26/05/2023	26/05/2023
13213	2023		47265	Dual-chamber implantable pacemaker, rate-responsive	CT141	Electrodes/Leads and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	16/05/2023	16/05/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	16/05/2023	16/05/2023
					CT145	Cardiovascular devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	16/05/2023	16/05/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	16/05/2023	16/05/2023
					CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	16/05/2023	16/05/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	16/05/2023	16/05/2023
					CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	16/05/2023	16/05/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	16/05/2023	16/05/2023
					CT233	Surgical	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	16/05/2023	16/05/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	16/05/2023	16/05/2023
					CT2406	Implantable	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	16/05/2023	16/05/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	16/05/2023	16/05/2023
					CT2407	Active implantable	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	16/05/2023	16/05/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	16/05/2023	16/05/2023

					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	16/05/2023	16/05/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	16/05/2023	16/05/2023
					CT2727	Active implant devices/ancillaries	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	16/05/2023	16/05/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	16/05/2023	16/05/2023
					CT2750	Implantable cardiac electrical stimulators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	16/05/2023	16/05/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	16/05/2023	16/05/2023
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	16/05/2023	16/05/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	16/05/2023	16/05/2023
					CT2985	Active	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	16/05/2023	16/05/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	16/05/2023	16/05/2023
					CT3001	Implantable metal	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	16/05/2023	16/05/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	16/05/2023	16/05/2023
					CT3002	Implantable synthetic polymer	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	16/05/2023	16/05/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	16/05/2023	16/05/2023
					CT3006	Implantable inorganic material	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	16/05/2023	16/05/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	16/05/2023	16/05/2023
					CT314	Electrically powered	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	16/05/2023	16/05/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	16/05/2023	16/05/2023
					CT321	Long-term surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	16/05/2023	16/05/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	16/05/2023	16/05/2023
					CT449	Generators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	16/05/2023	16/05/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	16/05/2023	16/05/2023
					CT981	Single use	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	16/05/2023	16/05/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	16/05/2023	16/05/2023
					CT983	Surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	16/05/2023	16/05/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	16/05/2023	16/05/2023
13214	2023	All other reportable incidents	47267	Single-chamber implantable pacemaker, rate-responsive	CT141	Electrodes/Leads and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	10/08/2022	10/08/2022
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	10/08/2022	10/08/2022
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	10/08/2022	10/08/2022
					CT145	Cardiovascular devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	10/08/2022	10/08/2022
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	10/08/2022	10/08/2022
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	10/08/2022	10/08/2022
					CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	10/08/2022	10/08/2022
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	10/08/2022	10/08/2022
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	10/08/2022	10/08/2022
					CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	10/08/2022	10/08/2022
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	10/08/2022	10/08/2022
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	10/08/2022	10/08/2022
					CT233	Surgical	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	10/08/2022	10/08/2022
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	10/08/2022	10/08/2022
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	10/08/2022	10/08/2022
					CT2406	Implantable	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	10/08/2022	10/08/2022
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	10/08/2022	10/08/2022
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	10/08/2022	10/08/2022
					CT2407	Active implantable	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	10/08/2022	10/08/2022
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	10/08/2022	10/08/2022
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	10/08/2022	10/08/2022
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	10/08/2022	10/08/2022
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	10/08/2022	10/08/2022
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	10/08/2022	10/08/2022
					CT2727	Active implant devices/ancillaries	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	10/08/2022	10/08/2022
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	10/08/2022	10/08/2022
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	10/08/2022	10/08/2022
					CT2750	Implantable cardiac electrical stimulators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	10/08/2022	10/08/2022
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	10/08/2022	10/08/2022
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	10/08/2022	10/08/2022
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	10/08/2022	10/08/2022
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	10/08/2022	10/08/2022
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	10/08/2022	10/08/2022
					CT2985	Active	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	10/08/2022	10/08/2022
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	10/08/2022	10/08/2022
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	10/08/2022	10/08/2022
					CT3001	Implantable metal	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	10/08/2022	10/08/2022
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	10/08/2022	10/08/2022
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	10/08/2022	10/08/2022
					CT3002	Implantable synthetic polymer	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	10/08/2022	10/08/2022
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	10/08/2022	10/08/2022
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	10/08/2022	10/08/2022
					CT3006	Implantable inorganic material	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	10/08/2022	10/08/2022
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	10/08/2022	10/08/2022
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	10/08/2022	10/08/2022
					CT314	Electrically powered	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	10/08/2022	10/08/2022
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	10/08/2022	10/08/2022
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	10/08/2022	10/08/2022
					CT321	Long-term surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	10/08/2022	10/08/2022
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	10/08/2022	10/08/2022
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	10/08/2022	10/08/2022
					CT449	Generators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	10/08/2022	10/08/2022
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	10/08/2022	10/08/2022
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	10/08/2022	10/08/2022
					CT981	Single use	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	10/08/2022	10/08/2022
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	10/08/2022	10/08/2022
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	10/08/2022	10/08/2022
					CT983	Surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	10/08/2022	10/08/2022
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	10/08/2022	10/08/2022
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	10/08/2022	10/08/2022
13215	2023		47265	Dual-chamber implantable pacemaker, rate-responsive	CT141	Electrodes/Leads and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	05/04/2023	05/04/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	05/04/2023	05/04/2023

					CT145	Cardiovascular devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	05/04/2023	05/04/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	05/04/2023	05/04/2023
					CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	05/04/2023	05/04/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	05/04/2023	05/04/2023
					CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	05/04/2023	05/04/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	05/04/2023	05/04/2023
					CT233	Surgical	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	05/04/2023	05/04/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	05/04/2023	05/04/2023
					CT2406	Implantable	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	05/04/2023	05/04/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	05/04/2023	05/04/2023
					CT2407	Active implantable	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	05/04/2023	05/04/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	05/04/2023	05/04/2023
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	05/04/2023	05/04/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	05/04/2023	05/04/2023
					CT2727	Active implant devices/ancillaries	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	05/04/2023	05/04/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	05/04/2023	05/04/2023
					CT2750	Implantable cardiac electrical stimulators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	05/04/2023	05/04/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	05/04/2023	05/04/2023
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	05/04/2023	05/04/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	05/04/2023	05/04/2023
					CT2985	Active	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	05/04/2023	05/04/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	05/04/2023	05/04/2023
					CT3001	Implantable metal	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	05/04/2023	05/04/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	05/04/2023	05/04/2023
					CT3002	Implantable synthetic polymer	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	05/04/2023	05/04/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	05/04/2023	05/04/2023
					CT3006	Implantable inorganic material	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	05/04/2023	05/04/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	05/04/2023	05/04/2023
					CT314	Electrically powered	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	05/04/2023	05/04/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	05/04/2023	05/04/2023
					CT321	Long-term surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	05/04/2023	05/04/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	05/04/2023	05/04/2023
					CT449	Generators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	05/04/2023	05/04/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	05/04/2023	05/04/2023
					CT981	Single use	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	05/04/2023	05/04/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	05/04/2023	05/04/2023
					CT983	Surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	05/04/2023	05/04/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	05/04/2023	05/04/2023
13216	2023	All other reportable incidents	47265	Dual-chamber implantable pacemaker, rate-responsive	CT141	Electrodes/Leads and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem		
							F1901	Additional Surgery	A24	No Apparent Device Problem		
					CT145	Cardiovascular devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem		
							F1901	Additional Surgery	A24	No Apparent Device Problem		
					CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem		
							F1901	Additional Surgery	A24	No Apparent Device Problem		
					CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem		
							F1901	Additional Surgery	A24	No Apparent Device Problem		
					CT233	Surgical	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem		
							F1901	Additional Surgery	A24	No Apparent Device Problem		
					CT2406	Implantable	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem		
							F1901	Additional Surgery	A24	No Apparent Device Problem		
					CT2407	Active implantable	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem		
							F1901	Additional Surgery	A24	No Apparent Device Problem		
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem		
							F1901	Additional Surgery	A24	No Apparent Device Problem		
					CT2727	Active implant devices/ancillaries	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem		
							F1901	Additional Surgery	A24	No Apparent Device Problem		
					CT2750	Implantable cardiac electrical stimulators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem		
							F1901	Additional Surgery	A24	No Apparent Device Problem		
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem		
							F1901	Additional Surgery	A24	No Apparent Device Problem		
					CT2985	Active	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem		
							F1901	Additional Surgery	A24	No Apparent Device Problem		
					CT3001	Implantable metal	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem		
							F1901	Additional Surgery	A24	No Apparent Device Problem		
					CT3002	Implantable synthetic polymer	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem		
							F1901	Additional Surgery	A24	No Apparent Device Problem		
					CT3006	Implantable inorganic material	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem		
							F1901	Additional Surgery	A24	No Apparent Device Problem		
					CT314	Electrically powered	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem		
							F1901	Additional Surgery	A24	No Apparent Device Problem		
					CT321	Long-term surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem		
							F1901	Additional Surgery	A24	No Apparent Device Problem		
					CT449	Generators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem		
							F1901	Additional Surgery	A24	No Apparent Device Problem		
					CT981	Single use	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem		
							F1901	Additional Surgery	A24	No Apparent Device Problem		
					CT983	Surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem		
							F1901	Additional Surgery	A24	No Apparent Device Problem		
13217	2023	All other reportable incidents	36241	Pacemaker/defibrillator lead	CT1684	Cardiac pacemakers and associated devices	E012207	Twitching	A051201	Device Dislodged or Dislocated		
									A070101	Failure to Capture		
									A071209	Extracardiac Stimulation		
							E2103	Device Overstimulation of Tissue	A051201	Device Dislodged or Dislocated		
									A070101	Failure to Capture		
									A071209	Extracardiac Stimulation		
							F08	Hospitalization or Prolonged Hospitalization	A051201	Device Dislodged or Dislocated		
									A070101	Failure to Capture		
									A071209	Extracardiac Stimulation		
							F1904	Device Repositioning	A051201	Device Dislodged or Dislocated		

								A070101	Failure to Capture			
								A071209	Extracardiac Stimulation			
						F23	Unexpected Medical Intervention	A051201	Device Dislodged or Dislocated			
								A070101	Failure to Capture			
								A071209	Extracardiac Stimulation			
13218	2023	All other reportable incidents	17882	Physiologic-monitoring defibrillation system	CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070803	Failure to Power Up		
							F27	Problem Identified During Non-Clinical Procedure	A070803	Failure to Power Up		
13219	2023	All other reportable incidents	47265	Dual-chamber implantable pacemaker, rate-responsive	CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	14/03/2023	14/03/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	14/03/2023	14/03/2023
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	14/03/2023	14/03/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	14/03/2023	14/03/2023
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	14/03/2023	14/03/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	14/03/2023	14/03/2023
13220	2023	All other reportable incidents	47265	Dual-chamber implantable pacemaker, rate-responsive	CT141	Electrodes/Leads and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F1901	Additional Surgery	A070504	Premature Discharge of Battery		
					CT145	Cardiovascular devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F1901	Additional Surgery	A070504	Premature Discharge of Battery		
					CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F1901	Additional Surgery	A070504	Premature Discharge of Battery		
					CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F1901	Additional Surgery	A070504	Premature Discharge of Battery		
					CT233	Surgical	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F1901	Additional Surgery	A070504	Premature Discharge of Battery		
					CT2406	Implantable	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F1901	Additional Surgery	A070504	Premature Discharge of Battery		
					CT2407	Active implantable	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F1901	Additional Surgery	A070504	Premature Discharge of Battery		
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F1901	Additional Surgery	A070504	Premature Discharge of Battery		
					CT2727	Active implant devices/ancillaries	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F1901	Additional Surgery	A070504	Premature Discharge of Battery		
					CT2750	Implantable cardiac electrical stimulators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F1901	Additional Surgery	A070504	Premature Discharge of Battery		
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F1901	Additional Surgery	A070504	Premature Discharge of Battery		
					CT2985	Active	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F1901	Additional Surgery	A070504	Premature Discharge of Battery		
					CT3001	Implantable metal	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F1901	Additional Surgery	A070504	Premature Discharge of Battery		
					CT3002	Implantable synthetic polymer	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F1901	Additional Surgery	A070504	Premature Discharge of Battery		
					CT3006	Implantable inorganic material	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F1901	Additional Surgery	A070504	Premature Discharge of Battery		
					CT314	Electrically powered	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F1901	Additional Surgery	A070504	Premature Discharge of Battery		
					CT321	Long-term surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F1901	Additional Surgery	A070504	Premature Discharge of Battery		
					CT449	Generators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F1901	Additional Surgery	A070504	Premature Discharge of Battery		
					CT981	Single use	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F1901	Additional Surgery	A070504	Premature Discharge of Battery		
					CT983	Surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F1901	Additional Surgery	A070504	Premature Discharge of Battery		
13221	2023	All other reportable incidents	35223	Endocardial/interventricular septal pacing lead	CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070104	Unstable Capture Threshold		
									A070909	Oversensing		
									A0722	Impedance Problem		
									A090801	Signal Artifact/Noise		
							F26	No Health Consequences or Impact	A070104	Unstable Capture Threshold		
									A070909	Oversensing		
									A0722	Impedance Problem		
									A090801	Signal Artifact/Noise		
13222	2023	All other reportable incidents	47205	Cardiac pulse generator programmer	CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A0902	Display or Visual Feedback Problem		
									A1104	Computer Operating System Problem		
							F26	No Health Consequences or Impact	A0902	Display or Visual Feedback Problem		
									A1104	Computer Operating System Problem		
13223	2023	All other reportable incidents	35223	Endocardial/interventricular septal pacing lead	CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A072201	High Impedance		
							F26	No Health Consequences or Impact	A072201	High Impedance		
13224	2023	All other reportable incidents	36241	Pacemaker/defibrillator lead	CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070909	Oversensing		
									A072201	High Impedance		
									A1301	Failure to Read Input Signal		
							F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A070909	Oversensing		
									A072201	High Impedance		
									A1301	Failure to Read Input Signal		
13225	2023	All other reportable incidents	47263	Cardiac resynchronization therapy implantable pacemaker	CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	15/11/2022	15/11/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	15/11/2022	15/11/2022
					CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	15/11/2022	15/11/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	15/11/2022	15/11/2022
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	15/11/2022	15/11/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	15/11/2022	15/11/2022
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	15/11/2022	15/11/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	15/11/2022	15/11/2022
13226	2023	All other reportable incidents	47263	Cardiac resynchronization therapy implantable pacemaker	CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	22/11/2022	22/11/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	22/11/2022	22/11/2022
					CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	22/11/2022	22/11/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	22/11/2022	22/11/2022
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	22/11/2022	22/11/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	22/11/2022	22/11/2022
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	22/11/2022	22/11/2022

							F1901	Additional Surgery	A24	No Apparent Device Problem	22/11/2022	22/11/2022
13227	2023	All other reportable incidents	47263	Cardiac resynchronization therapy implantable pacemaker	CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	24/11/2022	24/11/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	24/11/2022	24/11/2022
					CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	24/11/2022	24/11/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	24/11/2022	24/11/2022
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	24/11/2022	24/11/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	24/11/2022	24/11/2022
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	24/11/2022	24/11/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	24/11/2022	24/11/2022
13228	2023	All other reportable incidents	47205	Cardiac pulse generator programmer	CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A150105	Key or Button Unresponsive/not Working		
							F1908	Prolonged Surgery	A150105	Key or Button Unresponsive/not Working		
13229	2023	All other reportable incidents	47263	Cardiac resynchronization therapy implantable pacemaker	CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	29/11/2022	29/11/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	29/11/2022	29/11/2022
					CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	29/11/2022	29/11/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	29/11/2022	29/11/2022
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	29/11/2022	29/11/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	29/11/2022	29/11/2022
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	29/11/2022	29/11/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	29/11/2022	29/11/2022
13230	2023	All other reportable incidents	60789	Intracardiac pacemaker	CT1685	Cardiac pacemakers	E060106	Heart Block	A0701	Capturing Problem	24/05/2023	24/05/2023
									A070101	Failure to Capture	24/05/2023	24/05/2023
									A070102	High Capture Threshold	24/05/2023	24/05/2023
									A070104	Unstable Capture Threshold	24/05/2023	24/05/2023
							F08	Hospitalization or Prolonged Hospitalization	A0701	Capturing Problem	24/05/2023	24/05/2023
									A070101	Failure to Capture	24/05/2023	24/05/2023
									A070102	High Capture Threshold	24/05/2023	24/05/2023
									A070104	Unstable Capture Threshold	24/05/2023	24/05/2023
							F12	Serious Injury/ Illness/ Impairment	A0701	Capturing Problem	24/05/2023	24/05/2023
									A070101	Failure to Capture	24/05/2023	24/05/2023
									A070102	High Capture Threshold	24/05/2023	24/05/2023
									A070104	Unstable Capture Threshold	24/05/2023	24/05/2023
							F1905	Device Revision or Replacement	A0701	Capturing Problem	24/05/2023	24/05/2023
									A070101	Failure to Capture	24/05/2023	24/05/2023
									A070102	High Capture Threshold	24/05/2023	24/05/2023
									A070104	Unstable Capture Threshold	24/05/2023	24/05/2023
					CT257	Implantable cardiac pacemakers	E060106	Heart Block	A0701	Capturing Problem	24/05/2023	24/05/2023
									A070101	Failure to Capture	24/05/2023	24/05/2023
									A070102	High Capture Threshold	24/05/2023	24/05/2023
									A070104	Unstable Capture Threshold	24/05/2023	24/05/2023
							F08	Hospitalization or Prolonged Hospitalization	A0701	Capturing Problem	24/05/2023	24/05/2023
									A070101	Failure to Capture	24/05/2023	24/05/2023
									A070102	High Capture Threshold	24/05/2023	24/05/2023
									A070104	Unstable Capture Threshold	24/05/2023	24/05/2023
							F12	Serious Injury/ Illness/ Impairment	A0701	Capturing Problem	24/05/2023	24/05/2023
									A070101	Failure to Capture	24/05/2023	24/05/2023
									A070102	High Capture Threshold	24/05/2023	24/05/2023
									A070104	Unstable Capture Threshold	24/05/2023	24/05/2023
							F1905	Device Revision or Replacement	A0701	Capturing Problem	24/05/2023	24/05/2023
									A070101	Failure to Capture	24/05/2023	24/05/2023
									A070102	High Capture Threshold	24/05/2023	24/05/2023
									A070104	Unstable Capture Threshold	24/05/2023	24/05/2023
					CT2818	Catheterization/Cannulation/Fluid path devices	E060106	Heart Block	A0701	Capturing Problem	24/05/2023	24/05/2023
									A070101	Failure to Capture	24/05/2023	24/05/2023
									A070102	High Capture Threshold	24/05/2023	24/05/2023
									A070104	Unstable Capture Threshold	24/05/2023	24/05/2023
							F08	Hospitalization or Prolonged Hospitalization	A0701	Capturing Problem	24/05/2023	24/05/2023
									A070101	Failure to Capture	24/05/2023	24/05/2023
									A070102	High Capture Threshold	24/05/2023	24/05/2023
									A070104	Unstable Capture Threshold	24/05/2023	24/05/2023
							F12	Serious Injury/ Illness/ Impairment	A0701	Capturing Problem	24/05/2023	24/05/2023
									A070101	Failure to Capture	24/05/2023	24/05/2023
									A070102	High Capture Threshold	24/05/2023	24/05/2023
									A070104	Unstable Capture Threshold	24/05/2023	24/05/2023
							F1905	Device Revision or Replacement	A0701	Capturing Problem	24/05/2023	24/05/2023
									A070101	Failure to Capture	24/05/2023	24/05/2023
									A070102	High Capture Threshold	24/05/2023	24/05/2023
									A070104	Unstable Capture Threshold	24/05/2023	24/05/2023
					CT2866	Implantable cardiac electrical stimulators	E060106	Heart Block	A0701	Capturing Problem	24/05/2023	24/05/2023
									A070101	Failure to Capture	24/05/2023	24/05/2023
									A070102	High Capture Threshold	24/05/2023	24/05/2023
									A070104	Unstable Capture Threshold	24/05/2023	24/05/2023
							F08	Hospitalization or Prolonged Hospitalization	A0701	Capturing Problem	24/05/2023	24/05/2023
									A070101	Failure to Capture	24/05/2023	24/05/2023
									A070102	High Capture Threshold	24/05/2023	24/05/2023
									A070104	Unstable Capture Threshold	24/05/2023	24/05/2023
							F12	Serious Injury/ Illness/ Impairment	A0701	Capturing Problem	24/05/2023	24/05/2023
									A070101	Failure to Capture	24/05/2023	24/05/2023
									A070102	High Capture Threshold	24/05/2023	24/05/2023
									A070104	Unstable Capture Threshold	24/05/2023	24/05/2023
							F1905	Device Revision or Replacement	A0701	Capturing Problem	24/05/2023	24/05/2023
									A070101	Failure to Capture	24/05/2023	24/05/2023
									A070102	High Capture Threshold	24/05/2023	24/05/2023
									A070104	Unstable Capture Threshold	24/05/2023	24/05/2023
13231	2023	All other reportable incidents	47265	Dual-chamber implantable pacemaker, rate-responsive	CT141	Electrodes/Leads and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F1901	Additional Surgery	A070504	Premature Discharge of Battery		
					CT145	Cardiovascular devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F1901	Additional Surgery	A070504	Premature Discharge of Battery		
					CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		

							F1901	Additional Surgery	A070504	Premature Discharge of Battery		
					CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F1901	Additional Surgery	A070504	Premature Discharge of Battery		
					CT233	Surgical	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F1901	Additional Surgery	A070504	Premature Discharge of Battery		
					CT2406	Implantable	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F1901	Additional Surgery	A070504	Premature Discharge of Battery		
					CT2407	Active implantable	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F1901	Additional Surgery	A070504	Premature Discharge of Battery		
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F1901	Additional Surgery	A070504	Premature Discharge of Battery		
					CT2727	Active implant devices/ancillaries	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F1901	Additional Surgery	A070504	Premature Discharge of Battery		
					CT2750	Implantable cardiac electrical stimulators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F1901	Additional Surgery	A070504	Premature Discharge of Battery		
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F1901	Additional Surgery	A070504	Premature Discharge of Battery		
					CT2985	Active	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F1901	Additional Surgery	A070504	Premature Discharge of Battery		
					CT3001	Implantable metal	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F1901	Additional Surgery	A070504	Premature Discharge of Battery		
					CT3002	Implantable synthetic polymer	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F1901	Additional Surgery	A070504	Premature Discharge of Battery		
					CT3006	Implantable inorganic material	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F1901	Additional Surgery	A070504	Premature Discharge of Battery		
					CT314	Electrically powered	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F1901	Additional Surgery	A070504	Premature Discharge of Battery		
					CT321	Long-term surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F1901	Additional Surgery	A070504	Premature Discharge of Battery		
					CT449	Generators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F1901	Additional Surgery	A070504	Premature Discharge of Battery		
					CT981	Single use	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F1901	Additional Surgery	A070504	Premature Discharge of Battery		
					CT983	Surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F1901	Additional Surgery	A070504	Premature Discharge of Battery		
13232	2023	All other reportable incidents	36241	Pacemaker/defibrillator lead	CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	05/05/2023	05/05/2023
							F08	Hospitalization or Prolonged Hospitalization	A24	No Apparent Device Problem	05/05/2023	05/05/2023
							F1903	Device Explantation	A24	No Apparent Device Problem	05/05/2023	05/05/2023
13233	2023		47265	Dual-chamber implantable pacemaker, rate-responsive	CT141	Electrodes/Leads and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	15/05/2023	15/05/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	15/05/2023	15/05/2023
					CT145	Cardiovascular devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	15/05/2023	15/05/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	15/05/2023	15/05/2023
					CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	15/05/2023	15/05/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	15/05/2023	15/05/2023
					CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	15/05/2023	15/05/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	15/05/2023	15/05/2023
					CT233	Surgical	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	15/05/2023	15/05/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	15/05/2023	15/05/2023
					CT2406	Implantable	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	15/05/2023	15/05/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	15/05/2023	15/05/2023
					CT2407	Active implantable	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	15/05/2023	15/05/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	15/05/2023	15/05/2023
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	15/05/2023	15/05/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	15/05/2023	15/05/2023
					CT2727	Active implant devices/ancillaries	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	15/05/2023	15/05/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	15/05/2023	15/05/2023
					CT2750	Implantable cardiac electrical stimulators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	15/05/2023	15/05/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	15/05/2023	15/05/2023
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	15/05/2023	15/05/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	15/05/2023	15/05/2023
					CT2985	Active	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	15/05/2023	15/05/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	15/05/2023	15/05/2023
					CT3001	Implantable metal	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	15/05/2023	15/05/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	15/05/2023	15/05/2023
					CT3002	Implantable synthetic polymer	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	15/05/2023	15/05/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	15/05/2023	15/05/2023
					CT3006	Implantable inorganic material	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	15/05/2023	15/05/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	15/05/2023	15/05/2023
					CT314	Electrically powered	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	15/05/2023	15/05/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	15/05/2023	15/05/2023
					CT321	Long-term surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	15/05/2023	15/05/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	15/05/2023	15/05/2023
					CT449	Generators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	15/05/2023	15/05/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	15/05/2023	15/05/2023
					CT981	Single use	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	15/05/2023	15/05/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	15/05/2023	15/05/2023
					CT983	Surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	15/05/2023	15/05/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	15/05/2023	15/05/2023
13234	2023	All other reportable incidents	17882	Physiologic-monitoring defibrillation system	CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A0801	Failure to Calibrate		
							F27	Problem Identified During Non-Clinical Procedure	A0801	Failure to Calibrate		
13235	2023	All other reportable incidents	47267	Single-chamber implantable pacemaker, rate-responsive	CT141	Electrodes/Leads and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT145	Cardiovascular devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		

					CT233	Surgical	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT2406	Implantable	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT2407	Active implantable	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT2727	Active implant devices/ancillaries	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT2750	Implantable cardiac electrical stimulators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT2985	Active	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT3001	Implantable metal	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT3002	Implantable synthetic polymer	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT3006	Implantable inorganic material	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT314	Electrically powered	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT321	Long-term surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT449	Generators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT981	Single use	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT983	Surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
13236	2023	All other reportable incidents	47265	Dual-chamber implantable pacemaker, rate-responsive	CT141	Electrodes/Leads and associated devices	E0112	Dizziness	A070101	Failure to Capture		
							E011903	Syncope/Fainting	A070101	Failure to Capture		
							E060101	Asystole	A070101	Failure to Capture		
							E2007	Fall	A070101	Failure to Capture		
							F08	Hospitalization or Prolonged Hospitalization	A070101	Failure to Capture		
							F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A070101	Failure to Capture		
					CT145	Cardiovascular devices	E0112	Dizziness	A070101	Failure to Capture		
							E011903	Syncope/Fainting	A070101	Failure to Capture		
							E060101	Asystole	A070101	Failure to Capture		
							E2007	Fall	A070101	Failure to Capture		
							F08	Hospitalization or Prolonged Hospitalization	A070101	Failure to Capture		
							F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A070101	Failure to Capture		
					CT1684	Cardiac pacemakers and associated devices	E0112	Dizziness	A070101	Failure to Capture		
							E011903	Syncope/Fainting	A070101	Failure to Capture		
							E060101	Asystole	A070101	Failure to Capture		
							E2007	Fall	A070101	Failure to Capture		
							F08	Hospitalization or Prolonged Hospitalization	A070101	Failure to Capture		
							F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A070101	Failure to Capture		
					CT1685	Cardiac pacemakers	E0112	Dizziness	A070101	Failure to Capture		
							E011903	Syncope/Fainting	A070101	Failure to Capture		
							E060101	Asystole	A070101	Failure to Capture		
							E2007	Fall	A070101	Failure to Capture		
							F08	Hospitalization or Prolonged Hospitalization	A070101	Failure to Capture		
							F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A070101	Failure to Capture		
					CT233	Surgical	E0112	Dizziness	A070101	Failure to Capture		
							E011903	Syncope/Fainting	A070101	Failure to Capture		
							E060101	Asystole	A070101	Failure to Capture		
							E2007	Fall	A070101	Failure to Capture		
							F08	Hospitalization or Prolonged Hospitalization	A070101	Failure to Capture		
							F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A070101	Failure to Capture		
					CT2406	Implantable	E0112	Dizziness	A070101	Failure to Capture		
							E011903	Syncope/Fainting	A070101	Failure to Capture		
							E060101	Asystole	A070101	Failure to Capture		
							E2007	Fall	A070101	Failure to Capture		
							F08	Hospitalization or Prolonged Hospitalization	A070101	Failure to Capture		
							F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A070101	Failure to Capture		
					CT2407	Active implantable	E0112	Dizziness	A070101	Failure to Capture		
							E011903	Syncope/Fainting	A070101	Failure to Capture		
							E060101	Asystole	A070101	Failure to Capture		
							E2007	Fall	A070101	Failure to Capture		
							F08	Hospitalization or Prolonged Hospitalization	A070101	Failure to Capture		
							F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A070101	Failure to Capture		
					CT257	Implantable cardiac pacemakers	E0112	Dizziness	A070101	Failure to Capture		
							E011903	Syncope/Fainting	A070101	Failure to Capture		
							E060101	Asystole	A070101	Failure to Capture		
							E2007	Fall	A070101	Failure to Capture		
							F08	Hospitalization or Prolonged Hospitalization	A070101	Failure to Capture		
							F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A070101	Failure to Capture		
					CT2727	Active implant devices/ancillaries	E0112	Dizziness	A070101	Failure to Capture		
							E011903	Syncope/Fainting	A070101	Failure to Capture		
							E060101	Asystole	A070101	Failure to Capture		
							E2007	Fall	A070101	Failure to Capture		
							F08	Hospitalization or Prolonged Hospitalization	A070101	Failure to Capture		
							F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A070101	Failure to Capture		
					CT2750	Implantable cardiac electrical stimulators and associated devices	E0112	Dizziness	A070101	Failure to Capture		
							E011903	Syncope/Fainting	A070101	Failure to Capture		

							E060101	Asystole	A070101	Failure to Capture		
							E2007	Fall	A070101	Failure to Capture		
							F08	Hospitalization or Prolonged Hospitalization	A070101	Failure to Capture		
							F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A070101	Failure to Capture		
					CT2866	Implantable cardiac electrical stimulators	E0112	Dizziness	A070101	Failure to Capture		
							E011903	Syncope/Fainting	A070101	Failure to Capture		
							E060101	Asystole	A070101	Failure to Capture		
							E2007	Fall	A070101	Failure to Capture		
							F08	Hospitalization or Prolonged Hospitalization	A070101	Failure to Capture		
							F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A070101	Failure to Capture		
					CT2985	Active	E0112	Dizziness	A070101	Failure to Capture		
							E011903	Syncope/Fainting	A070101	Failure to Capture		
							E060101	Asystole	A070101	Failure to Capture		
							E2007	Fall	A070101	Failure to Capture		
							F08	Hospitalization or Prolonged Hospitalization	A070101	Failure to Capture		
							F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A070101	Failure to Capture		
					CT3001	Implantable metal	E0112	Dizziness	A070101	Failure to Capture		
							E011903	Syncope/Fainting	A070101	Failure to Capture		
							E060101	Asystole	A070101	Failure to Capture		
							E2007	Fall	A070101	Failure to Capture		
							F08	Hospitalization or Prolonged Hospitalization	A070101	Failure to Capture		
							F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A070101	Failure to Capture		
					CT3002	Implantable synthetic polymer	E0112	Dizziness	A070101	Failure to Capture		
							E011903	Syncope/Fainting	A070101	Failure to Capture		
							E060101	Asystole	A070101	Failure to Capture		
							E2007	Fall	A070101	Failure to Capture		
							F08	Hospitalization or Prolonged Hospitalization	A070101	Failure to Capture		
							F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A070101	Failure to Capture		
					CT3006	Implantable inorganic material	E0112	Dizziness	A070101	Failure to Capture		
							E011903	Syncope/Fainting	A070101	Failure to Capture		
							E060101	Asystole	A070101	Failure to Capture		
							E2007	Fall	A070101	Failure to Capture		
							F08	Hospitalization or Prolonged Hospitalization	A070101	Failure to Capture		
							F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A070101	Failure to Capture		
					CT314	Electrically powered	E0112	Dizziness	A070101	Failure to Capture		
							E011903	Syncope/Fainting	A070101	Failure to Capture		
							E060101	Asystole	A070101	Failure to Capture		
							E2007	Fall	A070101	Failure to Capture		
							F08	Hospitalization or Prolonged Hospitalization	A070101	Failure to Capture		
							F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A070101	Failure to Capture		
					CT321	Long-term surgical invasive	E0112	Dizziness	A070101	Failure to Capture		
							E011903	Syncope/Fainting	A070101	Failure to Capture		
							E060101	Asystole	A070101	Failure to Capture		
							E2007	Fall	A070101	Failure to Capture		
							F08	Hospitalization or Prolonged Hospitalization	A070101	Failure to Capture		
							F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A070101	Failure to Capture		
					CT449	Generators and associated devices	E0112	Dizziness	A070101	Failure to Capture		
							E011903	Syncope/Fainting	A070101	Failure to Capture		
							E060101	Asystole	A070101	Failure to Capture		
							E2007	Fall	A070101	Failure to Capture		
							F08	Hospitalization or Prolonged Hospitalization	A070101	Failure to Capture		
							F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A070101	Failure to Capture		
					CT981	Single use	E0112	Dizziness	A070101	Failure to Capture		
							E011903	Syncope/Fainting	A070101	Failure to Capture		
							E060101	Asystole	A070101	Failure to Capture		
							E2007	Fall	A070101	Failure to Capture		
							F08	Hospitalization or Prolonged Hospitalization	A070101	Failure to Capture		
							F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A070101	Failure to Capture		
					CT983	Surgical invasive	E0112	Dizziness	A070101	Failure to Capture		
							E011903	Syncope/Fainting	A070101	Failure to Capture		
							E060101	Asystole	A070101	Failure to Capture		
							E2007	Fall	A070101	Failure to Capture		
							F08	Hospitalization or Prolonged Hospitalization	A070101	Failure to Capture		
							F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A070101	Failure to Capture		
13237	2023	All other reportable incidents	60789	Intracardiac pacemaker	CT130	Catheters and associated devices	E060101	Asystole	A070101	Failure to Capture		
									A070504	Premature Discharge of Battery		
									A0712	Pacing Problem		
							F08	Hospitalization or Prolonged Hospitalization	A070101	Failure to Capture		
									A070504	Premature Discharge of Battery		
									A0712	Pacing Problem		
							F1203	Life Threatening Illness or Injury	A070101	Failure to Capture		
									A070504	Premature Discharge of Battery		
									A0712	Pacing Problem		
							F1905	Device Revision or Replacement	A070101	Failure to Capture		
									A070504	Premature Discharge of Battery		
									A0712	Pacing Problem		
					CT145	Cardiovascular devices	E060101	Asystole	A070101	Failure to Capture		
									A070504	Premature Discharge of Battery		
									A0712	Pacing Problem		
							F08	Hospitalization or Prolonged Hospitalization	A070101	Failure to Capture		
									A070504	Premature Discharge of Battery		
									A0712	Pacing Problem		
							F1203	Life Threatening Illness or Injury	A070101	Failure to Capture		
									A070504	Premature Discharge of Battery		
									A0712	Pacing Problem		
							F1905	Device Revision or Replacement	A070101	Failure to Capture		
									A070504	Premature Discharge of Battery		
									A0712	Pacing Problem		

					CT1684	Cardiac pacemakers and associated devices	E060101	Asystole	A070101	Failure to Capture
									A070504	Premature Discharge of Battery
									A0712	Pacing Problem
							F08	Hospitalization or Prolonged Hospitalization	A070101	Failure to Capture
									A070504	Premature Discharge of Battery
									A0712	Pacing Problem
							F1203	Life Threatening Illness or Injury	A070101	Failure to Capture
									A070504	Premature Discharge of Battery
									A0712	Pacing Problem
							F1905	Device Revision or Replacement	A070101	Failure to Capture
									A070504	Premature Discharge of Battery
									A0712	Pacing Problem
					CT1685	Cardiac pacemakers	E060101	Asystole	A070101	Failure to Capture
									A070504	Premature Discharge of Battery
									A0712	Pacing Problem
							F08	Hospitalization or Prolonged Hospitalization	A070101	Failure to Capture
									A070504	Premature Discharge of Battery
									A0712	Pacing Problem
							F1203	Life Threatening Illness or Injury	A070101	Failure to Capture
									A070504	Premature Discharge of Battery
									A0712	Pacing Problem
							F1905	Device Revision or Replacement	A070101	Failure to Capture
									A070504	Premature Discharge of Battery
									A0712	Pacing Problem
					CT233	Surgical	E060101	Asystole	A070101	Failure to Capture
									A070504	Premature Discharge of Battery
									A0712	Pacing Problem
							F08	Hospitalization or Prolonged Hospitalization	A070101	Failure to Capture
									A070504	Premature Discharge of Battery
									A0712	Pacing Problem
							F1203	Life Threatening Illness or Injury	A070101	Failure to Capture
									A070504	Premature Discharge of Battery
									A0712	Pacing Problem
							F1905	Device Revision or Replacement	A070101	Failure to Capture
									A070504	Premature Discharge of Battery
									A0712	Pacing Problem
					CT2406	Implantable	E060101	Asystole	A070101	Failure to Capture
									A070504	Premature Discharge of Battery
									A0712	Pacing Problem
							F08	Hospitalization or Prolonged Hospitalization	A070101	Failure to Capture
									A070504	Premature Discharge of Battery
									A0712	Pacing Problem
							F1203	Life Threatening Illness or Injury	A070101	Failure to Capture
									A070504	Premature Discharge of Battery
									A0712	Pacing Problem
							F1905	Device Revision or Replacement	A070101	Failure to Capture
									A070504	Premature Discharge of Battery
									A0712	Pacing Problem
					CT2407	Active implantable	E060101	Asystole	A070101	Failure to Capture
									A070504	Premature Discharge of Battery
									A0712	Pacing Problem
							F08	Hospitalization or Prolonged Hospitalization	A070101	Failure to Capture
									A070504	Premature Discharge of Battery
									A0712	Pacing Problem
							F1203	Life Threatening Illness or Injury	A070101	Failure to Capture
									A070504	Premature Discharge of Battery
									A0712	Pacing Problem
							F1905	Device Revision or Replacement	A070101	Failure to Capture
									A070504	Premature Discharge of Battery
									A0712	Pacing Problem
					CT257	Implantable cardiac pacemakers	E060101	Asystole	A070101	Failure to Capture
									A070504	Premature Discharge of Battery
									A0712	Pacing Problem
							F08	Hospitalization or Prolonged Hospitalization	A070101	Failure to Capture
									A070504	Premature Discharge of Battery
									A0712	Pacing Problem
							F1203	Life Threatening Illness or Injury	A070101	Failure to Capture
									A070504	Premature Discharge of Battery
									A0712	Pacing Problem
							F1905	Device Revision or Replacement	A070101	Failure to Capture
									A070504	Premature Discharge of Battery
									A0712	Pacing Problem
					CT2727	Active implant devices/ancillaries	E060101	Asystole	A070101	Failure to Capture
									A070504	Premature Discharge of Battery
									A0712	Pacing Problem
							F08	Hospitalization or Prolonged Hospitalization	A070101	Failure to Capture
									A070504	Premature Discharge of Battery
									A0712	Pacing Problem
							F1203	Life Threatening Illness or Injury	A070101	Failure to Capture
									A070504	Premature Discharge of Battery
									A0712	Pacing Problem
							F1905	Device Revision or Replacement	A070101	Failure to Capture
									A070504	Premature Discharge of Battery
									A0712	Pacing Problem
					CT2750	Implantable cardiac electrical stimulators and associated devices	E060101	Asystole	A070101	Failure to Capture
									A070504	Premature Discharge of Battery
									A0712	Pacing Problem
							F08	Hospitalization or Prolonged Hospitalization	A070101	Failure to Capture

									A070504	Premature Discharge of Battery		
									A0712	Pacing Problem		
							F1203	Life Threatening Illness or Injury	A070101	Failure to Capture		
									A070504	Premature Discharge of Battery		
									A0712	Pacing Problem		
							F1905	Device Revision or Replacement	A070101	Failure to Capture		
									A070504	Premature Discharge of Battery		
									A0712	Pacing Problem		
					CT2818	Catheterization/Cannulation/Fluid path devices	E060101	Asystole	A070101	Failure to Capture		
									A070504	Premature Discharge of Battery		
									A0712	Pacing Problem		
							F08	Hospitalization or Prolonged Hospitalization	A070101	Failure to Capture		
									A070504	Premature Discharge of Battery		
									A0712	Pacing Problem		
							F1203	Life Threatening Illness or Injury	A070101	Failure to Capture		
									A070504	Premature Discharge of Battery		
									A0712	Pacing Problem		
							F1905	Device Revision or Replacement	A070101	Failure to Capture		
									A070504	Premature Discharge of Battery		
									A0712	Pacing Problem		
					CT2866	Implantable cardiac electrical stimulators	E060101	Asystole	A070101	Failure to Capture		
									A070504	Premature Discharge of Battery		
									A0712	Pacing Problem		
							F08	Hospitalization or Prolonged Hospitalization	A070101	Failure to Capture		
									A070504	Premature Discharge of Battery		
									A0712	Pacing Problem		
							F1203	Life Threatening Illness or Injury	A070101	Failure to Capture		
									A070504	Premature Discharge of Battery		
									A0712	Pacing Problem		
							F1905	Device Revision or Replacement	A070101	Failure to Capture		
									A070504	Premature Discharge of Battery		
									A0712	Pacing Problem		
					CT2985	Active	E060101	Asystole	A070101	Failure to Capture		
									A070504	Premature Discharge of Battery		
									A0712	Pacing Problem		
							F08	Hospitalization or Prolonged Hospitalization	A070101	Failure to Capture		
									A070504	Premature Discharge of Battery		
									A0712	Pacing Problem		
							F1203	Life Threatening Illness or Injury	A070101	Failure to Capture		
									A070504	Premature Discharge of Battery		
									A0712	Pacing Problem		
							F1905	Device Revision or Replacement	A070101	Failure to Capture		
									A070504	Premature Discharge of Battery		
									A0712	Pacing Problem		
					CT3001	Implantable metal	E060101	Asystole	A070101	Failure to Capture		
									A070504	Premature Discharge of Battery		
									A0712	Pacing Problem		
							F08	Hospitalization or Prolonged Hospitalization	A070101	Failure to Capture		
									A070504	Premature Discharge of Battery		
									A0712	Pacing Problem		
							F1203	Life Threatening Illness or Injury	A070101	Failure to Capture		
									A070504	Premature Discharge of Battery		
									A0712	Pacing Problem		
							F1905	Device Revision or Replacement	A070101	Failure to Capture		
									A070504	Premature Discharge of Battery		
									A0712	Pacing Problem		
					CT3002	Implantable synthetic polymer	E060101	Asystole	A070101	Failure to Capture		
									A070504	Premature Discharge of Battery		
									A0712	Pacing Problem		
							F08	Hospitalization or Prolonged Hospitalization	A070101	Failure to Capture		
									A070504	Premature Discharge of Battery		
									A0712	Pacing Problem		
							F1203	Life Threatening Illness or Injury	A070101	Failure to Capture		
									A070504	Premature Discharge of Battery		
									A0712	Pacing Problem		
							F1905	Device Revision or Replacement	A070101	Failure to Capture		
									A070504	Premature Discharge of Battery		
									A0712	Pacing Problem		
					CT3006	Implantable inorganic material	E060101	Asystole	A070101	Failure to Capture		
									A070504	Premature Discharge of Battery		
									A0712	Pacing Problem		
							F08	Hospitalization or Prolonged Hospitalization	A070101	Failure to Capture		
									A070504	Premature Discharge of Battery		
									A0712	Pacing Problem		
							F1203	Life Threatening Illness or Injury	A070101	Failure to Capture		
									A070504	Premature Discharge of Battery		
									A0712	Pacing Problem		
							F1905	Device Revision or Replacement	A070101	Failure to Capture		
									A070504	Premature Discharge of Battery		
									A0712	Pacing Problem		
					CT314	Electrically powered	E060101	Asystole	A070101	Failure to Capture		
									A070504	Premature Discharge of Battery		
									A0712	Pacing Problem		
							F08	Hospitalization or Prolonged Hospitalization	A070101	Failure to Capture		
									A070504	Premature Discharge of Battery		
									A0712	Pacing Problem		
							F1203	Life Threatening Illness or Injury	A070101	Failure to Capture		
									A070504	Premature Discharge of Battery		

									A0712	Pacing Problem		
							F1905	Device Revision or Replacement	A070101	Failure to Capture		
									A070504	Premature Discharge of Battery		
									A0712	Pacing Problem		
					CT321	Long-term surgical invasive	E060101	Asystole	A070101	Failure to Capture		
									A070504	Premature Discharge of Battery		
									A0712	Pacing Problem		
							F08	Hospitalization or Prolonged Hospitalization	A070101	Failure to Capture		
									A070504	Premature Discharge of Battery		
									A0712	Pacing Problem		
							F1203	Life Threatening Illness or Injury	A070101	Failure to Capture		
									A070504	Premature Discharge of Battery		
									A0712	Pacing Problem		
							F1905	Device Revision or Replacement	A070101	Failure to Capture		
									A070504	Premature Discharge of Battery		
									A0712	Pacing Problem		
					CT981	Single use	E060101	Asystole	A070101	Failure to Capture		
									A070504	Premature Discharge of Battery		
									A0712	Pacing Problem		
							F08	Hospitalization or Prolonged Hospitalization	A070101	Failure to Capture		
									A070504	Premature Discharge of Battery		
									A0712	Pacing Problem		
							F1203	Life Threatening Illness or Injury	A070101	Failure to Capture		
									A070504	Premature Discharge of Battery		
									A0712	Pacing Problem		
							F1905	Device Revision or Replacement	A070101	Failure to Capture		
									A070504	Premature Discharge of Battery		
									A0712	Pacing Problem		
					CT983	Surgical invasive	E060101	Asystole	A070101	Failure to Capture		
									A070504	Premature Discharge of Battery		
									A0712	Pacing Problem		
							F08	Hospitalization or Prolonged Hospitalization	A070101	Failure to Capture		
									A070504	Premature Discharge of Battery		
									A0712	Pacing Problem		
							F1203	Life Threatening Illness or Injury	A070101	Failure to Capture		
									A070504	Premature Discharge of Battery		
									A0712	Pacing Problem		
							F1905	Device Revision or Replacement	A070101	Failure to Capture		
									A070504	Premature Discharge of Battery		
									A0712	Pacing Problem		
13238	2023	All other reportable incidents	35223	Endocardial/interventricular septal pacing lead	CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A0401	Break	19/05/2023	19/05/2023
									A070909	Oversensing	19/05/2023	19/05/2023
							F1901	Additional Surgery	A0401	Break	19/05/2023	19/05/2023
									A070909	Oversensing	19/05/2023	19/05/2023
13239	2023	All other reportable incidents	47205	Cardiac pulse generator programmer	CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A0203	Defective Device		
									A150105	Key or Button Unresponsive/not Working		
							F26	No Health Consequences or Impact	A0203	Defective Device		
									A150105	Key or Button Unresponsive/not Working		
13240	2023		35223	Endocardial/interventricular septal pacing lead	CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A0203	Defective Device		
									A04	Material Integrity Problem		
									A0709	Device Sensing Problem		
							F1905	Device Revision or Replacement	A0203	Defective Device		
									A04	Material Integrity Problem		
									A0709	Device Sensing Problem		
13241	2023		47265	Dual-chamber implantable pacemaker, rate-responsive	CT141	Electrodes/Leads and associated devices	F26	No Health Consequences or Impact	A0712	Pacing Problem		
									A09	Output Problem		
					CT145	Cardiovascular devices	F26	No Health Consequences or Impact	A0712	Pacing Problem		
									A09	Output Problem		
					CT1684	Cardiac pacemakers and associated devices	F26	No Health Consequences or Impact	A0712	Pacing Problem		
									A09	Output Problem		
					CT1685	Cardiac pacemakers	F26	No Health Consequences or Impact	A0712	Pacing Problem		
									A09	Output Problem		
					CT233	Surgical	F26	No Health Consequences or Impact	A0712	Pacing Problem		
									A09	Output Problem		
					CT2406	Implantable	F26	No Health Consequences or Impact	A0712	Pacing Problem		
									A09	Output Problem		
					CT2407	Active implantable	F26	No Health Consequences or Impact	A0712	Pacing Problem		
									A09	Output Problem		
					CT257	Implantable cardiac pacemakers	F26	No Health Consequences or Impact	A0712	Pacing Problem		
									A09	Output Problem		
					CT2727	Active implant devices/ancillaries	F26	No Health Consequences or Impact	A0712	Pacing Problem		
									A09	Output Problem		
					CT2750	Implantable cardiac electrical stimulators and associated devices	F26	No Health Consequences or Impact	A0712	Pacing Problem		
									A09	Output Problem		
					CT2866	Implantable cardiac electrical stimulators	F26	No Health Consequences or Impact	A0712	Pacing Problem		
									A09	Output Problem		
					CT2985	Active	F26	No Health Consequences or Impact	A0712	Pacing Problem		
									A09	Output Problem		
					CT3001	Implantable metal	F26	No Health Consequences or Impact	A0712	Pacing Problem		
									A09	Output Problem		
					CT3002	Implantable synthetic polymer	F26	No Health Consequences or Impact	A0712	Pacing Problem		
									A09	Output Problem		
					CT3006	Implantable inorganic material	F26	No Health Consequences or Impact	A0712	Pacing Problem		
									A09	Output Problem		
					CT314	Electrically powered	F26	No Health Consequences or Impact	A0712	Pacing Problem		
									A09	Output Problem		
					CT321	Long-term surgical invasive	F26	No Health Consequences or Impact	A0712	Pacing Problem		
									A09	Output Problem		

					CT449	Generators and associated devices	F26	No Health Consequences or Impact	A0712	Pacing Problem		
									A09	Output Problem		
					CT981	Single use	F26	No Health Consequences or Impact	A0712	Pacing Problem		
									A09	Output Problem		
					CT983	Surgical invasive	F26	No Health Consequences or Impact	A0712	Pacing Problem		
									A09	Output Problem		
13242	2023	All other reportable incidents	35852	Single-chamber implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
13243	2023	All other reportable incidents	47205	Cardiac pulse generator programmer	CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A150105	Key or Button Unresponsive/not Working		
							F1908	Prolonged Surgery	A150105	Key or Button Unresponsive/not Working		
13244	2023	All other reportable incidents	37265	Dual-chamber implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	30/05/2023	30/05/2023
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	30/05/2023	30/05/2023
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	30/05/2023	30/05/2023
13245	2023	All other reportable incidents	47267	Single-chamber implantable pacemaker, rate-responsive	CT141	Electrodes/Leads and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	02/11/2022	02/11/2022
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	02/11/2022	02/11/2022
							F1903	Device Explantation	A070504	Premature Discharge of Battery	02/11/2022	02/11/2022
					CT145	Cardiovascular devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	02/11/2022	02/11/2022
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	02/11/2022	02/11/2022
							F1903	Device Explantation	A070504	Premature Discharge of Battery	02/11/2022	02/11/2022
					CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	02/11/2022	02/11/2022
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	02/11/2022	02/11/2022
							F1903	Device Explantation	A070504	Premature Discharge of Battery	02/11/2022	02/11/2022
					CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	02/11/2022	02/11/2022
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	02/11/2022	02/11/2022
							F1903	Device Explantation	A070504	Premature Discharge of Battery	02/11/2022	02/11/2022
					CT233	Surgical	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	02/11/2022	02/11/2022
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	02/11/2022	02/11/2022
							F1903	Device Explantation	A070504	Premature Discharge of Battery	02/11/2022	02/11/2022
					CT2406	Implantable	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	02/11/2022	02/11/2022
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	02/11/2022	02/11/2022
							F1903	Device Explantation	A070504	Premature Discharge of Battery	02/11/2022	02/11/2022
					CT2407	Active implantable	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	02/11/2022	02/11/2022
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	02/11/2022	02/11/2022
							F1903	Device Explantation	A070504	Premature Discharge of Battery	02/11/2022	02/11/2022
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	02/11/2022	02/11/2022
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	02/11/2022	02/11/2022
							F1903	Device Explantation	A070504	Premature Discharge of Battery	02/11/2022	02/11/2022
					CT2727	Active implant devices/ancillaries	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	02/11/2022	02/11/2022
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	02/11/2022	02/11/2022
							F1903	Device Explantation	A070504	Premature Discharge of Battery	02/11/2022	02/11/2022
					CT2750	Implantable cardiac electrical stimulators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	02/11/2022	02/11/2022
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	02/11/2022	02/11/2022
							F1903	Device Explantation	A070504	Premature Discharge of Battery	02/11/2022	02/11/2022
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	02/11/2022	02/11/2022
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	02/11/2022	02/11/2022
							F1903	Device Explantation	A070504	Premature Discharge of Battery	02/11/2022	02/11/2022
					CT2985	Active	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	02/11/2022	02/11/2022
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	02/11/2022	02/11/2022
							F1903	Device Explantation	A070504	Premature Discharge of Battery	02/11/2022	02/11/2022
					CT3001	Implantable metal	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	02/11/2022	02/11/2022
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	02/11/2022	02/11/2022
							F1903	Device Explantation	A070504	Premature Discharge of Battery	02/11/2022	02/11/2022
					CT3002	Implantable synthetic polymer	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	02/11/2022	02/11/2022
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	02/11/2022	02/11/2022
							F1903	Device Explantation	A070504	Premature Discharge of Battery	02/11/2022	02/11/2022
					CT3006	Implantable inorganic material	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	02/11/2022	02/11/2022
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	02/11/2022	02/11/2022
							F1903	Device Explantation	A070504	Premature Discharge of Battery	02/11/2022	02/11/2022
					CT314	Electrically powered	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	02/11/2022	02/11/2022
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	02/11/2022	02/11/2022
							F1903	Device Explantation	A070504	Premature Discharge of Battery	02/11/2022	02/11/2022
					CT321	Long-term surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	02/11/2022	02/11/2022
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	02/11/2022	02/11/2022
							F1903	Device Explantation	A070504	Premature Discharge of Battery	02/11/2022	02/11/2022
					CT449	Generators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	02/11/2022	02/11/2022
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	02/11/2022	02/11/2022
							F1903	Device Explantation	A070504	Premature Discharge of Battery	02/11/2022	02/11/2022
					CT981	Single use	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	02/11/2022	02/11/2022
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	02/11/2022	02/11/2022
							F1903	Device Explantation	A070504	Premature Discharge of Battery	02/11/2022	02/11/2022
					CT983	Surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	02/11/2022	02/11/2022
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	02/11/2022	02/11/2022
							F1903	Device Explantation	A070504	Premature Discharge of Battery	02/11/2022	02/11/2022
13246	2023	All other reportable incidents	47265	Dual-chamber implantable pacemaker, rate-responsive	CT141	Electrodes/Leads and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	19/05/2023	19/05/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	19/05/2023	19/05/2023
					CT145	Cardiovascular devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	19/05/2023	19/05/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	19/05/2023	19/05/2023
					CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	19/05/2023	19/05/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	19/05/2023	19/05/2023
					CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	19/05/2023	19/05/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	19/05/2023	19/05/2023
					CT233	Surgical	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	19/05/2023	19/05/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	19/05/2023	19/05/2023
					CT2406	Implantable	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	19/05/2023	19/05/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	19/05/2023	19/05/2023
					CT2407	Active implantable	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	19/05/2023	19/05/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	19/05/2023	19/05/2023
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	19/05/2023	19/05/2023

						F1901	Additional Surgery	A24	No Apparent Device Problem	19/05/2023	19/05/2023
					CT2727	Active implant devices/ancillaries	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	19/05/2023	19/05/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	19/05/2023	19/05/2023
					CT2750	Implantable cardiac electrical stimulators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	19/05/2023	19/05/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	19/05/2023	19/05/2023
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	19/05/2023	19/05/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	19/05/2023	19/05/2023
					CT2985	Active	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	19/05/2023	19/05/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	19/05/2023	19/05/2023
					CT3001	Implantable metal	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	19/05/2023	19/05/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	19/05/2023	19/05/2023
					CT3002	Implantable synthetic polymer	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	19/05/2023	19/05/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	19/05/2023	19/05/2023
					CT3006	Implantable inorganic material	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	19/05/2023	19/05/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	19/05/2023	19/05/2023
					CT314	Electrically powered	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	19/05/2023	19/05/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	19/05/2023	19/05/2023
					CT321	Long-term surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	19/05/2023	19/05/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	19/05/2023	19/05/2023
					CT449	Generators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	19/05/2023	19/05/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	19/05/2023	19/05/2023
					CT981	Single use	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	19/05/2023	19/05/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	19/05/2023	19/05/2023
					CT983	Surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	19/05/2023	19/05/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	19/05/2023	19/05/2023
13247	2023	All other reportable incidents	36241	Pacemaker/defibrillator lead	CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A072201	High Impedance		
							F26	No Health Consequences or Impact	A072201	High Impedance		
13248	2023	All other reportable incidents	17882	Physiologic-monitoring defibrillation system	CT1684	Cardiac pacemakers and associated devices	E060110	Ventricular Fibrillation	A0714	Unintended Electrical Shock		
							F24	Insufficient Health Impact Information	A0714	Unintended Electrical Shock		
13249	2023	All other reportable incidents	35223	Endocardial/interventricular septal pacing lead	CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070102	High Capture Threshold	26/05/2023	26/05/2023
									A072201	High Impedance	26/05/2023	26/05/2023
							F08	Hospitalization or Prolonged Hospitalization	A070102	High Capture Threshold	26/05/2023	26/05/2023
									A072201	High Impedance	26/05/2023	26/05/2023
							F12	Serious Injury/ Illness/ Impairment	A070102	High Capture Threshold	26/05/2023	26/05/2023
									A072201	High Impedance	26/05/2023	26/05/2023
							F1905	Device Revision or Replacement	A070102	High Capture Threshold	26/05/2023	26/05/2023
									A072201	High Impedance	26/05/2023	26/05/2023
							F23	Unexpected Medical Intervention	A070102	High Capture Threshold	26/05/2023	26/05/2023
									A072201	High Impedance	26/05/2023	26/05/2023
13250	2023	All other reportable incidents	17882	Physiologic-monitoring defibrillation system	CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070803	Failure to Power Up		
							F27	Problem Identified During Non-Clinical Procedure	A070803	Failure to Power Up		
13251	2023		47265	Dual-chamber implantable pacemaker, rate-responsive	CT141	Electrodes/Leads and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	19/05/2023	19/05/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	19/05/2023	19/05/2023
					CT145	Cardiovascular devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	19/05/2023	19/05/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	19/05/2023	19/05/2023
					CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	19/05/2023	19/05/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	19/05/2023	19/05/2023
					CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	19/05/2023	19/05/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	19/05/2023	19/05/2023
					CT233	Surgical	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	19/05/2023	19/05/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	19/05/2023	19/05/2023
					CT2406	Implantable	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	19/05/2023	19/05/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	19/05/2023	19/05/2023
					CT2407	Active implantable	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	19/05/2023	19/05/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	19/05/2023	19/05/2023
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	19/05/2023	19/05/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	19/05/2023	19/05/2023
					CT2727	Active implant devices/ancillaries	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	19/05/2023	19/05/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	19/05/2023	19/05/2023
					CT2750	Implantable cardiac electrical stimulators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	19/05/2023	19/05/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	19/05/2023	19/05/2023
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	19/05/2023	19/05/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	19/05/2023	19/05/2023
					CT2985	Active	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	19/05/2023	19/05/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	19/05/2023	19/05/2023
					CT3001	Implantable metal	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	19/05/2023	19/05/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	19/05/2023	19/05/2023
					CT3002	Implantable synthetic polymer	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	19/05/2023	19/05/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	19/05/2023	19/05/2023
					CT3006	Implantable inorganic material	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	19/05/2023	19/05/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	19/05/2023	19/05/2023
					CT314	Electrically powered	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	19/05/2023	19/05/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	19/05/2023	19/05/2023
					CT321	Long-term surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	19/05/2023	19/05/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	19/05/2023	19/05/2023
					CT449	Generators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	19/05/2023	19/05/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	19/05/2023	19/05/2023
					CT981	Single use	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	19/05/2023	19/05/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	19/05/2023	19/05/2023
					CT983	Surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	19/05/2023	19/05/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	19/05/2023	19/05/2023
13252	2023	All other reportable incidents	47265	Dual-chamber implantable pacemaker, rate-responsive	CT141	Electrodes/Leads and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	18/05/2023	18/05/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	18/05/2023	18/05/2023
					CT145	Cardiovascular devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	18/05/2023	18/05/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	18/05/2023	18/05/2023
					CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	18/05/2023	18/05/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	18/05/2023	18/05/2023
					CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	18/05/2023	18/05/2023

							F1901	Additional Surgery	A24	No Apparent Device Problem	18/05/2023	18/05/2023
					CT233	Surgical	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	18/05/2023	18/05/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	18/05/2023	18/05/2023
					CT2406	Implantable	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	18/05/2023	18/05/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	18/05/2023	18/05/2023
					CT2407	Active implantable	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	18/05/2023	18/05/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	18/05/2023	18/05/2023
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	18/05/2023	18/05/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	18/05/2023	18/05/2023
					CT2727	Active implant devices/ancillaries	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	18/05/2023	18/05/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	18/05/2023	18/05/2023
					CT2750	Implantable cardiac electrical stimulators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	18/05/2023	18/05/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	18/05/2023	18/05/2023
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	18/05/2023	18/05/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	18/05/2023	18/05/2023
					CT2985	Active	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	18/05/2023	18/05/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	18/05/2023	18/05/2023
					CT3001	Implantable metal	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	18/05/2023	18/05/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	18/05/2023	18/05/2023
					CT3002	Implantable synthetic polymer	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	18/05/2023	18/05/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	18/05/2023	18/05/2023
					CT3006	Implantable inorganic material	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	18/05/2023	18/05/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	18/05/2023	18/05/2023
					CT314	Electrically powered	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	18/05/2023	18/05/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	18/05/2023	18/05/2023
					CT321	Long-term surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	18/05/2023	18/05/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	18/05/2023	18/05/2023
					CT449	Generators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	18/05/2023	18/05/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	18/05/2023	18/05/2023
					CT981	Single use	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	18/05/2023	18/05/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	18/05/2023	18/05/2023
					CT983	Surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	18/05/2023	18/05/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	18/05/2023	18/05/2023
13253	2023	All other reportable incidents	47263	Cardiac resynchronization therapy implantable pacemaker	CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	15/09/2022	15/09/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	15/09/2022	15/09/2022
					CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	15/09/2022	15/09/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	15/09/2022	15/09/2022
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	15/09/2022	15/09/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	15/09/2022	15/09/2022
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	15/09/2022	15/09/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	15/09/2022	15/09/2022
13254	2023	All other reportable incidents	47263	Cardiac resynchronization therapy implantable pacemaker	CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	20/09/2022	20/09/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	20/09/2022	20/09/2022
					CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	20/09/2022	20/09/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	20/09/2022	20/09/2022
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	20/09/2022	20/09/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	20/09/2022	20/09/2022
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	20/09/2022	20/09/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	20/09/2022	20/09/2022
13255	2023	All other reportable incidents	47263	Cardiac resynchronization therapy implantable pacemaker	CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	29/09/2022	29/09/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	29/09/2022	29/09/2022
					CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	29/09/2022	29/09/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	29/09/2022	29/09/2022
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	29/09/2022	29/09/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	29/09/2022	29/09/2022
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	29/09/2022	29/09/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	29/09/2022	29/09/2022
13256	2023	All other reportable incidents	47263	Cardiac resynchronization therapy implantable pacemaker	CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	11/10/2022	11/10/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	11/10/2022	11/10/2022
					CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	11/10/2022	11/10/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	11/10/2022	11/10/2022
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	11/10/2022	11/10/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	11/10/2022	11/10/2022
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	11/10/2022	11/10/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	11/10/2022	11/10/2022
13257	2023	All other reportable incidents	35852	Single-chamber implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem	20/04/2023	20/04/2023
									A070504	Premature Discharge of Battery	20/04/2023	20/04/2023
									A0908	Incorrect, Inadequate or Imprecise Result or Readings	20/04/2023	20/04/2023
							F08	Hospitalization or Prolonged Hospitalization	A0705	Battery Problem	20/04/2023	20/04/2023
									A070504	Premature Discharge of Battery	20/04/2023	20/04/2023
									A0908	Incorrect, Inadequate or Imprecise Result or Readings	20/04/2023	20/04/2023
							F1905	Device Revision or Replacement	A0705	Battery Problem	20/04/2023	20/04/2023
									A070504	Premature Discharge of Battery	20/04/2023	20/04/2023
									A0908	Incorrect, Inadequate or Imprecise Result or Readings	20/04/2023	20/04/2023
13258	2023	All other reportable incidents	47263	Cardiac resynchronization therapy implantable pacemaker	CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	06/09/2022	06/09/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	06/09/2022	06/09/2022
					CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	06/09/2022	06/09/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	06/09/2022	06/09/2022
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	06/09/2022	06/09/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	06/09/2022	06/09/2022
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	06/09/2022	06/09/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	06/09/2022	06/09/2022
13259	2023	All other reportable incidents	47263	Cardiac resynchronization therapy implantable pacemaker	CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	02/03/2023	02/03/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	02/03/2023	02/03/2023
					CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	02/03/2023	02/03/2023

							F1901	Additional Surgery	A24	No Apparent Device Problem	02/03/2023	02/03/2023
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	02/03/2023	02/03/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	02/03/2023	02/03/2023
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	02/03/2023	02/03/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	02/03/2023	02/03/2023
13260	2023	All other reportable incidents	47263	Cardiac resynchronization therapy implantable pacemaker	CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	06/09/2022	06/09/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	06/09/2022	06/09/2022
					CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	06/09/2022	06/09/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	06/09/2022	06/09/2022
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	06/09/2022	06/09/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	06/09/2022	06/09/2022
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	06/09/2022	06/09/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	06/09/2022	06/09/2022
13261	2023	All other reportable incidents	47263	Cardiac resynchronization therapy implantable pacemaker	CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	15/09/2022	15/09/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	15/09/2022	15/09/2022
					CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	15/09/2022	15/09/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	15/09/2022	15/09/2022
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	15/09/2022	15/09/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	15/09/2022	15/09/2022
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	15/09/2022	15/09/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	15/09/2022	15/09/2022
13262	2023	All other reportable incidents	47263	Cardiac resynchronization therapy implantable pacemaker	CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	07/02/2023	07/02/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	07/02/2023	07/02/2023
					CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	07/02/2023	07/02/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	07/02/2023	07/02/2023
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	07/02/2023	07/02/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	07/02/2023	07/02/2023
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	07/02/2023	07/02/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	07/02/2023	07/02/2023
13263	2023	All other reportable incidents	47265	Dual-chamber implantable pacemaker, rate-responsive	CT141	Electrodes/Leads and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
					CT145	Cardiovascular devices	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
					CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
					CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
					CT233	Surgical	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
					CT2406	Implantable	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
					CT2407	Active implantable	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
					CT2727	Active implant devices/ancillaries	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
					CT2750	Implantable cardiac electrical stimulators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
					CT2985	Active	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
					CT3001	Implantable metal	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
					CT3002	Implantable synthetic polymer	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
					CT3006	Implantable inorganic material	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A0705	Battery Problem		

									A070504	Premature Discharge of Battery		
					CT314	Electrically powered	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
					CT321	Long-term surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
					CT449	Generators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
					CT981	Single use	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
					CT983	Surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
13264	2023	All other reportable incidents	17882	Physiologic-monitoring defibrillation system	CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070803	Failure to Power Up		
							F27	Problem Identified During Non-Clinical Procedure	A070803	Failure to Power Up		
13265	2023	All other reportable incidents	47270	Cardiac resynchronization therapy implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A070603	Failure to Charge	06/06/2023	06/06/2023
							F1901	Additional Surgery	A070603	Failure to Charge	06/06/2023	06/06/2023
13266	2023	All other reportable incidents	47263	Cardiac resynchronization therapy implantable pacemaker	CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	21/03/2023	21/03/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	21/03/2023	21/03/2023
					CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	21/03/2023	21/03/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	21/03/2023	21/03/2023
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	21/03/2023	21/03/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	21/03/2023	21/03/2023
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	21/03/2023	21/03/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	21/03/2023	21/03/2023
13267	2023	All other reportable incidents	47263	Cardiac resynchronization therapy implantable pacemaker	CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	01/11/2022	01/11/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	01/11/2022	01/11/2022
					CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	01/11/2022	01/11/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	01/11/2022	01/11/2022
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	01/11/2022	01/11/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	01/11/2022	01/11/2022
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	01/11/2022	01/11/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	01/11/2022	01/11/2022
13268	2023	All other reportable incidents	47263	Cardiac resynchronization therapy implantable pacemaker	CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	18/10/2022	18/10/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	18/10/2022	18/10/2022
					CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	18/10/2022	18/10/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	18/10/2022	18/10/2022
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	18/10/2022	18/10/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	18/10/2022	18/10/2022
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	18/10/2022	18/10/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	18/10/2022	18/10/2022
13269	2023	All other reportable incidents	47263	Cardiac resynchronization therapy implantable pacemaker	CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	18/10/2022	18/10/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	18/10/2022	18/10/2022
					CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	18/10/2022	18/10/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	18/10/2022	18/10/2022
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	18/10/2022	18/10/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	18/10/2022	18/10/2022
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	18/10/2022	18/10/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	18/10/2022	18/10/2022
13270	2023	All other reportable incidents	63188	Left-ventricular wireless pacing system sensor/transmitter	CT145	Cardiovascular devices			A070504	Premature Discharge of Battery		
					CT1577	Wireless monitoring system transmitters			A070504	Premature Discharge of Battery		
					CT1684	Cardiac pacemakers and associated devices			A070504	Premature Discharge of Battery		
					CT2727	Active implant devices/ancillaries			A070504	Premature Discharge of Battery		
					CT2750	Implantable cardiac electrical stimulators and associated devices			A070504	Premature Discharge of Battery		
					CT2885	Physiological monitoring/measuring system transmitters/receivers/transceivers			A070504	Premature Discharge of Battery		
13271	2023	All other reportable incidents	47267	Single-chamber implantable pacemaker, rate-responsive	CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem	14/07/2023	14/07/2023
									A070504	Premature Discharge of Battery	14/07/2023	14/07/2023
							F08	Hospitalization or Prolonged Hospitalization	A0705	Battery Problem	14/07/2023	14/07/2023
									A070504	Premature Discharge of Battery	14/07/2023	14/07/2023
							F1905	Device Revision or Replacement	A0705	Battery Problem	14/07/2023	14/07/2023
									A070504	Premature Discharge of Battery	14/07/2023	14/07/2023
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem	14/07/2023	14/07/2023
									A070504	Premature Discharge of Battery	14/07/2023	14/07/2023
							F08	Hospitalization or Prolonged Hospitalization	A0705	Battery Problem	14/07/2023	14/07/2023
									A070504	Premature Discharge of Battery	14/07/2023	14/07/2023
							F1905	Device Revision or Replacement	A0705	Battery Problem	14/07/2023	14/07/2023
									A070504	Premature Discharge of Battery	14/07/2023	14/07/2023
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem	14/07/2023	14/07/2023
									A070504	Premature Discharge of Battery	14/07/2023	14/07/2023
							F08	Hospitalization or Prolonged Hospitalization	A0705	Battery Problem	14/07/2023	14/07/2023
									A070504	Premature Discharge of Battery	14/07/2023	14/07/2023
							F1905	Device Revision or Replacement	A0705	Battery Problem	14/07/2023	14/07/2023
									A070504	Premature Discharge of Battery	14/07/2023	14/07/2023
13272	2023	All other reportable incidents	47267	Single-chamber implantable pacemaker, rate-responsive	CT141	Electrodes/Leads and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	31/05/2023	31/05/2023
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	31/05/2023	31/05/2023
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	31/05/2023	31/05/2023
					CT145	Cardiovascular devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	31/05/2023	31/05/2023
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	31/05/2023	31/05/2023

							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	31/05/2023	31/05/2023
					CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	31/05/2023	31/05/2023
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	31/05/2023	31/05/2023
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	31/05/2023	31/05/2023
					CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	31/05/2023	31/05/2023
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	31/05/2023	31/05/2023
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	31/05/2023	31/05/2023
					CT233	Surgical	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	31/05/2023	31/05/2023
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	31/05/2023	31/05/2023
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	31/05/2023	31/05/2023
					CT2406	Implantable	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	31/05/2023	31/05/2023
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	31/05/2023	31/05/2023
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	31/05/2023	31/05/2023
					CT2407	Active implantable	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	31/05/2023	31/05/2023
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	31/05/2023	31/05/2023
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	31/05/2023	31/05/2023
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	31/05/2023	31/05/2023
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	31/05/2023	31/05/2023
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	31/05/2023	31/05/2023
					CT2727	Active implant devices/ancillaries	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	31/05/2023	31/05/2023
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	31/05/2023	31/05/2023
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	31/05/2023	31/05/2023
					CT2750	Implantable cardiac electrical stimulators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	31/05/2023	31/05/2023
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	31/05/2023	31/05/2023
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	31/05/2023	31/05/2023
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	31/05/2023	31/05/2023
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	31/05/2023	31/05/2023
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	31/05/2023	31/05/2023
					CT2985	Active	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	31/05/2023	31/05/2023
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	31/05/2023	31/05/2023
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	31/05/2023	31/05/2023
					CT3001	Implantable metal	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	31/05/2023	31/05/2023
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	31/05/2023	31/05/2023
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	31/05/2023	31/05/2023
					CT3002	Implantable synthetic polymer	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	31/05/2023	31/05/2023
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	31/05/2023	31/05/2023
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	31/05/2023	31/05/2023
					CT3006	Implantable inorganic material	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	31/05/2023	31/05/2023
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	31/05/2023	31/05/2023
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	31/05/2023	31/05/2023
					CT314	Electrically powered	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	31/05/2023	31/05/2023
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	31/05/2023	31/05/2023
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	31/05/2023	31/05/2023
					CT321	Long-term surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	31/05/2023	31/05/2023
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	31/05/2023	31/05/2023
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	31/05/2023	31/05/2023
					CT449	Generators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	31/05/2023	31/05/2023
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	31/05/2023	31/05/2023
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	31/05/2023	31/05/2023
					CT981	Single use	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	31/05/2023	31/05/2023
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	31/05/2023	31/05/2023
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	31/05/2023	31/05/2023
					CT983	Surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	31/05/2023	31/05/2023
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	31/05/2023	31/05/2023
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	31/05/2023	31/05/2023
13273	2023	All other reportable incidents	47265	Dual-chamber implantable pacemaker, rate-responsive	CT141	Electrodes/Leads and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
					CT145	Cardiovascular devices	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
					CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
					CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
					CT233	Surgical	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
					CT2406	Implantable	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
					CT2407	Active implantable	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
					CT2727	Active implant devices/ancillaries	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem		

									A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
					CT2750	Implantable cardiac electrical stimulators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
					CT2985	Active	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
					CT3001	Implantable metal	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
					CT3002	Implantable synthetic polymer	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
					CT3006	Implantable inorganic material	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
					CT314	Electrically powered	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
					CT321	Long-term surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
					CT449	Generators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
					CT981	Single use	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
					CT983	Surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
13274	2023	All other reportable incidents	47270	Cardiac resynchronization therapy implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
13275	2023	All other reportable incidents	47265	Dual-chamber implantable pacemaker, rate-responsive	CT141	Electrodes/Leads and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A071102	Failure to Interrogate	24/05/2023	24/05/2023
							F1901	Additional Surgery	A071102	Failure to Interrogate	24/05/2023	24/05/2023
					CT145	Cardiovascular devices	E2403	No Clinical Signs, Symptoms or Conditions	A071102	Failure to Interrogate	24/05/2023	24/05/2023
							F1901	Additional Surgery	A071102	Failure to Interrogate	24/05/2023	24/05/2023
					CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A071102	Failure to Interrogate	24/05/2023	24/05/2023
							F1901	Additional Surgery	A071102	Failure to Interrogate	24/05/2023	24/05/2023
					CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A071102	Failure to Interrogate	24/05/2023	24/05/2023
							F1901	Additional Surgery	A071102	Failure to Interrogate	24/05/2023	24/05/2023
					CT233	Surgical	E2403	No Clinical Signs, Symptoms or Conditions	A071102	Failure to Interrogate	24/05/2023	24/05/2023
							F1901	Additional Surgery	A071102	Failure to Interrogate	24/05/2023	24/05/2023
					CT2406	Implantable	E2403	No Clinical Signs, Symptoms or Conditions	A071102	Failure to Interrogate	24/05/2023	24/05/2023
							F1901	Additional Surgery	A071102	Failure to Interrogate	24/05/2023	24/05/2023
					CT2407	Active implantable	E2403	No Clinical Signs, Symptoms or Conditions	A071102	Failure to Interrogate	24/05/2023	24/05/2023
							F1901	Additional Surgery	A071102	Failure to Interrogate	24/05/2023	24/05/2023
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A071102	Failure to Interrogate	24/05/2023	24/05/2023
							F1901	Additional Surgery	A071102	Failure to Interrogate	24/05/2023	24/05/2023
					CT2727	Active implant devices/ancillaries	E2403	No Clinical Signs, Symptoms or Conditions	A071102	Failure to Interrogate	24/05/2023	24/05/2023
							F1901	Additional Surgery	A071102	Failure to Interrogate	24/05/2023	24/05/2023
					CT2750	Implantable cardiac electrical stimulators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A071102	Failure to Interrogate	24/05/2023	24/05/2023
							F1901	Additional Surgery	A071102	Failure to Interrogate	24/05/2023	24/05/2023
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A071102	Failure to Interrogate	24/05/2023	24/05/2023
							F1901	Additional Surgery	A071102	Failure to Interrogate	24/05/2023	24/05/2023
					CT2985	Active	E2403	No Clinical Signs, Symptoms or Conditions	A071102	Failure to Interrogate	24/05/2023	24/05/2023
							F1901	Additional Surgery	A071102	Failure to Interrogate	24/05/2023	24/05/2023
					CT3001	Implantable metal	E2403	No Clinical Signs, Symptoms or Conditions	A071102	Failure to Interrogate	24/05/2023	24/05/2023
							F1901	Additional Surgery	A071102	Failure to Interrogate	24/05/2023	24/05/2023
					CT3002	Implantable synthetic polymer	E2403	No Clinical Signs, Symptoms or Conditions	A071102	Failure to Interrogate	24/05/2023	24/05/2023
							F1901	Additional Surgery	A071102	Failure to Interrogate	24/05/2023	24/05/2023
					CT3006	Implantable inorganic material	E2403	No Clinical Signs, Symptoms or Conditions	A071102	Failure to Interrogate	24/05/2023	24/05/2023
							F1901	Additional Surgery	A071102	Failure to Interrogate	24/05/2023	24/05/2023
					CT314	Electrically powered	E2403	No Clinical Signs, Symptoms or Conditions	A071102	Failure to Interrogate	24/05/2023	24/05/2023
							F1901	Additional Surgery	A071102	Failure to Interrogate	24/05/2023	24/05/2023
					CT321	Long-term surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A071102	Failure to Interrogate	24/05/2023	24/05/2023
							F1901	Additional Surgery	A071102	Failure to Interrogate	24/05/2023	24/05/2023
					CT449	Generators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A071102	Failure to Interrogate	24/05/2023	24/05/2023
							F1901	Additional Surgery	A071102	Failure to Interrogate	24/05/2023	24/05/2023
					CT981	Single use	E2403	No Clinical Signs, Symptoms or Conditions	A071102	Failure to Interrogate	24/05/2023	24/05/2023
							F1901	Additional Surgery	A071102	Failure to Interrogate	24/05/2023	24/05/2023
					CT983	Surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A071102	Failure to Interrogate	24/05/2023	24/05/2023

							F1901	Additional Surgery	A071102	Failure to Interrogate	24/05/2023	24/05/2023
13276	2023	All other reportable incidents	47265	Dual-chamber implantable pacemaker, rate-responsive	CT141	Electrodes/Leads and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	19/05/2023	19/05/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	19/05/2023	19/05/2023
					CT145	Cardiovascular devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	19/05/2023	19/05/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	19/05/2023	19/05/2023
					CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	19/05/2023	19/05/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	19/05/2023	19/05/2023
					CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	19/05/2023	19/05/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	19/05/2023	19/05/2023
					CT233	Surgical	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	19/05/2023	19/05/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	19/05/2023	19/05/2023
					CT2406	Implantable	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	19/05/2023	19/05/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	19/05/2023	19/05/2023
					CT2407	Active implantable	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	19/05/2023	19/05/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	19/05/2023	19/05/2023
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	19/05/2023	19/05/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	19/05/2023	19/05/2023
					CT2727	Active implant devices/ancillaries	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	19/05/2023	19/05/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	19/05/2023	19/05/2023
					CT2750	Implantable cardiac electrical stimulators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	19/05/2023	19/05/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	19/05/2023	19/05/2023
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	19/05/2023	19/05/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	19/05/2023	19/05/2023
					CT2985	Active	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	19/05/2023	19/05/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	19/05/2023	19/05/2023
					CT3001	Implantable metal	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	19/05/2023	19/05/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	19/05/2023	19/05/2023
					CT3002	Implantable synthetic polymer	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	19/05/2023	19/05/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	19/05/2023	19/05/2023
					CT3006	Implantable inorganic material	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	19/05/2023	19/05/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	19/05/2023	19/05/2023
					CT314	Electrically powered	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	19/05/2023	19/05/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	19/05/2023	19/05/2023
					CT321	Long-term surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	19/05/2023	19/05/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	19/05/2023	19/05/2023
					CT449	Generators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	19/05/2023	19/05/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	19/05/2023	19/05/2023
					CT981	Single use	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	19/05/2023	19/05/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	19/05/2023	19/05/2023
					CT983	Surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	19/05/2023	19/05/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	19/05/2023	19/05/2023
13277	2023	All other reportable incidents	47265	Dual-chamber implantable pacemaker, rate-responsive	CT141	Electrodes/Leads and associated devices	E0601	Arrhythmia	A0711	Interrogation Problem	25/05/2023	25/05/2023
							F1901	Additional Surgery	A0711	Interrogation Problem	25/05/2023	25/05/2023
					CT145	Cardiovascular devices	E0601	Arrhythmia	A0711	Interrogation Problem	25/05/2023	25/05/2023
							F1901	Additional Surgery	A0711	Interrogation Problem	25/05/2023	25/05/2023
					CT1684	Cardiac pacemakers and associated devices	E0601	Arrhythmia	A0711	Interrogation Problem	25/05/2023	25/05/2023
							F1901	Additional Surgery	A0711	Interrogation Problem	25/05/2023	25/05/2023
					CT1685	Cardiac pacemakers	E0601	Arrhythmia	A0711	Interrogation Problem	25/05/2023	25/05/2023
							F1901	Additional Surgery	A0711	Interrogation Problem	25/05/2023	25/05/2023
					CT233	Surgical	E0601	Arrhythmia	A0711	Interrogation Problem	25/05/2023	25/05/2023
							F1901	Additional Surgery	A0711	Interrogation Problem	25/05/2023	25/05/2023
					CT2406	Implantable	E0601	Arrhythmia	A0711	Interrogation Problem	25/05/2023	25/05/2023
							F1901	Additional Surgery	A0711	Interrogation Problem	25/05/2023	25/05/2023
					CT2407	Active implantable	E0601	Arrhythmia	A0711	Interrogation Problem	25/05/2023	25/05/2023
							F1901	Additional Surgery	A0711	Interrogation Problem	25/05/2023	25/05/2023
					CT257	Implantable cardiac pacemakers	E0601	Arrhythmia	A0711	Interrogation Problem	25/05/2023	25/05/2023
							F1901	Additional Surgery	A0711	Interrogation Problem	25/05/2023	25/05/2023
					CT2727	Active implant devices/ancillaries	E0601	Arrhythmia	A0711	Interrogation Problem	25/05/2023	25/05/2023
							F1901	Additional Surgery	A0711	Interrogation Problem	25/05/2023	25/05/2023
					CT2750	Implantable cardiac electrical stimulators and associated devices	E0601	Arrhythmia	A0711	Interrogation Problem	25/05/2023	25/05/2023
							F1901	Additional Surgery	A0711	Interrogation Problem	25/05/2023	25/05/2023
					CT2866	Implantable cardiac electrical stimulators	E0601	Arrhythmia	A0711	Interrogation Problem	25/05/2023	25/05/2023
							F1901	Additional Surgery	A0711	Interrogation Problem	25/05/2023	25/05/2023
					CT2985	Active	E0601	Arrhythmia	A0711	Interrogation Problem	25/05/2023	25/05/2023
							F1901	Additional Surgery	A0711	Interrogation Problem	25/05/2023	25/05/2023
					CT3001	Implantable metal	E0601	Arrhythmia	A0711	Interrogation Problem	25/05/2023	25/05/2023
							F1901	Additional Surgery	A0711	Interrogation Problem	25/05/2023	25/05/2023
					CT3002	Implantable synthetic polymer	E0601	Arrhythmia	A0711	Interrogation Problem	25/05/2023	25/05/2023
							F1901	Additional Surgery	A0711	Interrogation Problem	25/05/2023	25/05/2023
					CT3006	Implantable inorganic material	E0601	Arrhythmia	A0711	Interrogation Problem	25/05/2023	25/05/2023
							F1901	Additional Surgery	A0711	Interrogation Problem	25/05/2023	25/05/2023
					CT314	Electrically powered	E0601	Arrhythmia	A0711	Interrogation Problem	25/05/2023	25/05/2023
							F1901	Additional Surgery	A0711	Interrogation Problem	25/05/2023	25/05/2023
					CT321	Long-term surgical invasive	E0601	Arrhythmia	A0711	Interrogation Problem	25/05/2023	25/05/2023
							F1901	Additional Surgery	A0711	Interrogation Problem	25/05/2023	25/05/2023
					CT449	Generators and associated devices	E0601	Arrhythmia	A0711	Interrogation Problem	25/05/2023	25/05/2023
							F1901	Additional Surgery	A0711	Interrogation Problem	25/05/2023	25/05/2023
					CT981	Single use	E0601	Arrhythmia	A0711	Interrogation Problem	25/05/2023	25/05/2023
							F1901	Additional Surgery	A0711	Interrogation Problem	25/05/2023	25/05/2023
					CT983	Surgical invasive	E0601	Arrhythmia	A0711	Interrogation Problem	25/05/2023	25/05/2023
							F1901	Additional Surgery	A0711	Interrogation Problem	25/05/2023	25/05/2023
13278	2023	All other reportable incidents	17882	Physiologic-monitoring defibrillation system	CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A1208	Fitting Problem		
							F26	No Health Consequences or Impact	A1208	Fitting Problem		
13279	2023	All other reportable incidents	36241	Pacemaker/defibrillator lead	CT1684	Cardiac pacemakers and associated devices	E2104	Electric Shock	A0203	Defective Device		
									A070909	Oversensing		
									A071302	Inappropriate/Inadequate Shock/Stimulation		
									A090801	Signal Artifact/Noise		
							F08	Hospitalization or Prolonged Hospitalization	A0203	Defective Device		

								A070909	Oversensing		
								A071302	Inappropriate/Inadequate Shock/Stimulation		
								A090801	Signal Artifact/Noise		
						F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A0203	Defective Device		
								A070909	Oversensing		
								A071302	Inappropriate/Inadequate Shock/Stimulation		
								A090801	Signal Artifact/Noise		
13280	2023	All other reportable incidents	47205	Cardiac pulse generator programmer	CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A150105	Key or Button Unresponsive/not Working		
							F1908	Prolonged Surgery	A150105	Key or Button Unresponsive/not Working		
13281	2023	All other reportable incidents	36241	Pacemaker/defibrillator lead	CT1684	Cardiac pacemakers and associated devices	E0131	Speech Disorder	A010402	Migration		
									A04	Material Integrity Problem		
									A040101	Fracture		
									A070101	Failure to Capture		
									A090801	Signal Artifact/Noise		
							E013302	Ischemia Stroke	A010402	Migration		
									A04	Material Integrity Problem		
									A040101	Fracture		
									A070101	Failure to Capture		
									A090801	Signal Artifact/Noise		
							E013403	Paresthesia	A010402	Migration		
									A04	Material Integrity Problem		
									A040101	Fracture		
									A070101	Failure to Capture		
									A090801	Signal Artifact/Noise		
							E0137	Transient Ischemic Attack	A010402	Migration		
									A04	Material Integrity Problem		
									A040101	Fracture		
									A070101	Failure to Capture		
									A090801	Signal Artifact/Noise		
							E050304	Thromboembolism	A010402	Migration		
									A04	Material Integrity Problem		
									A040101	Fracture		
									A070101	Failure to Capture		
									A090801	Signal Artifact/Noise		
							E083903	Diplopia/Double Vision	A010402	Migration		
									A04	Material Integrity Problem		
									A040101	Fracture		
									A070101	Failure to Capture		
									A090801	Signal Artifact/Noise		
							F08	Hospitalization or Prolonged Hospitalization	A010402	Migration		
									A04	Material Integrity Problem		
									A040101	Fracture		
									A070101	Failure to Capture		
									A090801	Signal Artifact/Noise		
							F1203	Life Threatening Illness or Injury	A010402	Migration		
									A04	Material Integrity Problem		
									A040101	Fracture		
									A070101	Failure to Capture		
									A090801	Signal Artifact/Noise		
							F1204	Permanent Impairment	A010402	Migration		
									A04	Material Integrity Problem		
									A040101	Fracture		
									A070101	Failure to Capture		
									A090801	Signal Artifact/Noise		
							F1905	Device Revision or Replacement	A010402	Migration		
									A04	Material Integrity Problem		
									A040101	Fracture		
									A070101	Failure to Capture		
									A090801	Signal Artifact/Noise		
							F2203	Imaging Required	A010402	Migration		
									A04	Material Integrity Problem		
									A040101	Fracture		
									A070101	Failure to Capture		
									A090801	Signal Artifact/Noise		
							F2303	Medication Required	A010402	Migration		
									A04	Material Integrity Problem		
									A040101	Fracture		
									A070101	Failure to Capture		
									A090801	Signal Artifact/Noise		
13282	2023	All other reportable incidents	47263	Cardiac resynchronization therapy implantable pacemaker	CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	10/03/2023	10/03/2023
							F1901	Additional Surgery	A070504	Premature Discharge of Battery	10/03/2023	10/03/2023
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	10/03/2023	10/03/2023
							F1901	Additional Surgery	A070504	Premature Discharge of Battery	10/03/2023	10/03/2023
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	10/03/2023	10/03/2023
							F1901	Additional Surgery	A070504	Premature Discharge of Battery	10/03/2023	10/03/2023
13283	2023	All other reportable incidents	47265	Dual-chamber implantable pacemaker, rate-responsive	CT141	Electrodes/Leads and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A072202	Low Impedance		
									A090801	Signal Artifact/Noise		
							F1908	Prolonged Surgery	A072202	Low Impedance		
									A090801	Signal Artifact/Noise		
					CT145	Cardiovascular devices	E2403	No Clinical Signs, Symptoms or Conditions	A072202	Low Impedance		
									A090801	Signal Artifact/Noise		
							F1908	Prolonged Surgery	A072202	Low Impedance		
									A090801	Signal Artifact/Noise		
					CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A072202	Low Impedance		
									A090801	Signal Artifact/Noise		
							F1908	Prolonged Surgery	A072202	Low Impedance		
									A090801	Signal Artifact/Noise		
					CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A072202	Low Impedance		

								A090801	Signal Artifact/Noise		
						F1908	Prolonged Surgery	A072202	Low Impedance		
								A090801	Signal Artifact/Noise		
					CT233	Surgical	E2403	No Clinical Signs, Symptoms or Conditions	A072202	Low Impedance		
									A090801	Signal Artifact/Noise		
						F1908	Prolonged Surgery	A072202	Low Impedance		
									A090801	Signal Artifact/Noise		
					CT2406	Implantable	E2403	No Clinical Signs, Symptoms or Conditions	A072202	Low Impedance		
									A090801	Signal Artifact/Noise		
						F1908	Prolonged Surgery	A072202	Low Impedance		
									A090801	Signal Artifact/Noise		
					CT2407	Active implantable	E2403	No Clinical Signs, Symptoms or Conditions	A072202	Low Impedance		
									A090801	Signal Artifact/Noise		
						F1908	Prolonged Surgery	A072202	Low Impedance		
									A090801	Signal Artifact/Noise		
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A072202	Low Impedance		
									A090801	Signal Artifact/Noise		
						F1908	Prolonged Surgery	A072202	Low Impedance		
									A090801	Signal Artifact/Noise		
					CT2727	Active implant devices/ancillaries	E2403	No Clinical Signs, Symptoms or Conditions	A072202	Low Impedance		
									A090801	Signal Artifact/Noise		
						F1908	Prolonged Surgery	A072202	Low Impedance		
									A090801	Signal Artifact/Noise		
					CT2750	Implantable cardiac electrical stimulators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A072202	Low Impedance		
									A090801	Signal Artifact/Noise		
						F1908	Prolonged Surgery	A072202	Low Impedance		
									A090801	Signal Artifact/Noise		
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A072202	Low Impedance		
									A090801	Signal Artifact/Noise		
						F1908	Prolonged Surgery	A072202	Low Impedance		
									A090801	Signal Artifact/Noise		
					CT2985	Active	E2403	No Clinical Signs, Symptoms or Conditions	A072202	Low Impedance		
									A090801	Signal Artifact/Noise		
						F1908	Prolonged Surgery	A072202	Low Impedance		
									A090801	Signal Artifact/Noise		
					CT3001	Implantable metal	E2403	No Clinical Signs, Symptoms or Conditions	A072202	Low Impedance		
									A090801	Signal Artifact/Noise		
						F1908	Prolonged Surgery	A072202	Low Impedance		
									A090801	Signal Artifact/Noise		
					CT3002	Implantable synthetic polymer	E2403	No Clinical Signs, Symptoms or Conditions	A072202	Low Impedance		
									A090801	Signal Artifact/Noise		
						F1908	Prolonged Surgery	A072202	Low Impedance		
									A090801	Signal Artifact/Noise		
					CT3006	Implantable inorganic material	E2403	No Clinical Signs, Symptoms or Conditions	A072202	Low Impedance		
									A090801	Signal Artifact/Noise		
						F1908	Prolonged Surgery	A072202	Low Impedance		
									A090801	Signal Artifact/Noise		
					CT314	Electrically powered	E2403	No Clinical Signs, Symptoms or Conditions	A072202	Low Impedance		
									A090801	Signal Artifact/Noise		
						F1908	Prolonged Surgery	A072202	Low Impedance		
									A090801	Signal Artifact/Noise		
					CT321	Long-term surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A072202	Low Impedance		
									A090801	Signal Artifact/Noise		
						F1908	Prolonged Surgery	A072202	Low Impedance		
									A090801	Signal Artifact/Noise		
					CT449	Generators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A072202	Low Impedance		
									A090801	Signal Artifact/Noise		
						F1908	Prolonged Surgery	A072202	Low Impedance		
									A090801	Signal Artifact/Noise		
					CT981	Single use	E2403	No Clinical Signs, Symptoms or Conditions	A072202	Low Impedance		
									A090801	Signal Artifact/Noise		
						F1908	Prolonged Surgery	A072202	Low Impedance		
									A090801	Signal Artifact/Noise		
					CT983	Surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A072202	Low Impedance		
									A090801	Signal Artifact/Noise		
						F1908	Prolonged Surgery	A072202	Low Impedance			
									A090801	Signal Artifact/Noise		
13284	2023	All other reportable incidents	47265	Dual-chamber implantable pacemaker, rate-responsive	CT141	Electrodes/Leads and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A072202	Low Impedance	20/02/2023	20/02/2023
									A090801	Signal Artifact/Noise	20/02/2023	20/02/2023
							F1905	Device Revision or Replacement	A072202	Low Impedance	20/02/2023	20/02/2023
									A090801	Signal Artifact/Noise	20/02/2023	20/02/2023
							F1908	Prolonged Surgery	A072202	Low Impedance	20/02/2023	20/02/2023
									A090801	Signal Artifact/Noise	20/02/2023	20/02/2023
					CT145	Cardiovascular devices	E2403	No Clinical Signs, Symptoms or Conditions	A072202	Low Impedance	20/02/2023	20/02/2023
									A090801	Signal Artifact/Noise	20/02/2023	20/02/2023
							F1905	Device Revision or Replacement	A072202	Low Impedance	20/02/2023	20/02/2023
									A090801	Signal Artifact/Noise	20/02/2023	20/02/2023
							F1908	Prolonged Surgery	A072202	Low Impedance	20/02/2023	20/02/2023
									A090801	Signal Artifact/Noise	20/02/2023	20/02/2023
					CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A072202	Low Impedance	20/02/2023	20/02/2023
									A090801	Signal Artifact/Noise	20/02/2023	20/02/2023
							F1905	Device Revision or Replacement	A072202	Low Impedance	20/02/2023	20/02/2023
									A090801	Signal Artifact/Noise	20/02/2023	20/02/2023
							F1908	Prolonged Surgery	A072202	Low Impedance	20/02/2023	20/02/2023
									A090801	Signal Artifact/Noise	20/02/2023	20/02/2023
					CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A072202	Low Impedance	20/02/2023	20/02/2023
									A090801	Signal Artifact/Noise	20/02/2023	20/02/2023
							F1905	Device Revision or Replacement	A072202	Low Impedance	20/02/2023	20/02/2023

							A090801	Signal Artifact/Noise	20/02/2023	20/02/2023
					F1908	Prolonged Surgery	A072202	Low Impedance	20/02/2023	20/02/2023
							A090801	Signal Artifact/Noise	20/02/2023	20/02/2023
				CT233	Surgical	E2403	No Clinical Signs, Symptoms or Conditions	A072202	Low Impedance	20/02/2023	20/02/2023
								A090801	Signal Artifact/Noise	20/02/2023	20/02/2023
						F1905	Device Revision or Replacement	A072202	Low Impedance	20/02/2023	20/02/2023
								A090801	Signal Artifact/Noise	20/02/2023	20/02/2023
						F1908	Prolonged Surgery	A072202	Low Impedance	20/02/2023	20/02/2023
								A090801	Signal Artifact/Noise	20/02/2023	20/02/2023
				CT2406	Implantable	E2403	No Clinical Signs, Symptoms or Conditions	A072202	Low Impedance	20/02/2023	20/02/2023
								A090801	Signal Artifact/Noise	20/02/2023	20/02/2023
						F1905	Device Revision or Replacement	A072202	Low Impedance	20/02/2023	20/02/2023
								A090801	Signal Artifact/Noise	20/02/2023	20/02/2023
						F1908	Prolonged Surgery	A072202	Low Impedance	20/02/2023	20/02/2023
								A090801	Signal Artifact/Noise	20/02/2023	20/02/2023
				CT2407	Active implantable	E2403	No Clinical Signs, Symptoms or Conditions	A072202	Low Impedance	20/02/2023	20/02/2023
								A090801	Signal Artifact/Noise	20/02/2023	20/02/2023
						F1905	Device Revision or Replacement	A072202	Low Impedance	20/02/2023	20/02/2023
								A090801	Signal Artifact/Noise	20/02/2023	20/02/2023
						F1908	Prolonged Surgery	A072202	Low Impedance	20/02/2023	20/02/2023
								A090801	Signal Artifact/Noise	20/02/2023	20/02/2023
				CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A072202	Low Impedance	20/02/2023	20/02/2023
								A090801	Signal Artifact/Noise	20/02/2023	20/02/2023
						F1905	Device Revision or Replacement	A072202	Low Impedance	20/02/2023	20/02/2023
								A090801	Signal Artifact/Noise	20/02/2023	20/02/2023
						F1908	Prolonged Surgery	A072202	Low Impedance	20/02/2023	20/02/2023
								A090801	Signal Artifact/Noise	20/02/2023	20/02/2023
				CT2727	Active implant devices/ancillaries	E2403	No Clinical Signs, Symptoms or Conditions	A072202	Low Impedance	20/02/2023	20/02/2023
								A090801	Signal Artifact/Noise	20/02/2023	20/02/2023
						F1905	Device Revision or Replacement	A072202	Low Impedance	20/02/2023	20/02/2023
								A090801	Signal Artifact/Noise	20/02/2023	20/02/2023
						F1908	Prolonged Surgery	A072202	Low Impedance	20/02/2023	20/02/2023
								A090801	Signal Artifact/Noise	20/02/2023	20/02/2023
				CT2750	Implantable cardiac electrical stimulators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A072202	Low Impedance	20/02/2023	20/02/2023
								A090801	Signal Artifact/Noise	20/02/2023	20/02/2023
						F1905	Device Revision or Replacement	A072202	Low Impedance	20/02/2023	20/02/2023
								A090801	Signal Artifact/Noise	20/02/2023	20/02/2023
						F1908	Prolonged Surgery	A072202	Low Impedance	20/02/2023	20/02/2023
								A090801	Signal Artifact/Noise	20/02/2023	20/02/2023
				CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A072202	Low Impedance	20/02/2023	20/02/2023
								A090801	Signal Artifact/Noise	20/02/2023	20/02/2023
						F1905	Device Revision or Replacement	A072202	Low Impedance	20/02/2023	20/02/2023
								A090801	Signal Artifact/Noise	20/02/2023	20/02/2023
						F1908	Prolonged Surgery	A072202	Low Impedance	20/02/2023	20/02/2023
								A090801	Signal Artifact/Noise	20/02/2023	20/02/2023
				CT2985	Active	E2403	No Clinical Signs, Symptoms or Conditions	A072202	Low Impedance	20/02/2023	20/02/2023
								A090801	Signal Artifact/Noise	20/02/2023	20/02/2023
						F1905	Device Revision or Replacement	A072202	Low Impedance	20/02/2023	20/02/2023
								A090801	Signal Artifact/Noise	20/02/2023	20/02/2023
						F1908	Prolonged Surgery	A072202	Low Impedance	20/02/2023	20/02/2023
								A090801	Signal Artifact/Noise	20/02/2023	20/02/2023
				CT3001	Implantable metal	E2403	No Clinical Signs, Symptoms or Conditions	A072202	Low Impedance	20/02/2023	20/02/2023
								A090801	Signal Artifact/Noise	20/02/2023	20/02/2023
						F1905	Device Revision or Replacement	A072202	Low Impedance	20/02/2023	20/02/2023
								A090801	Signal Artifact/Noise	20/02/2023	20/02/2023
						F1908	Prolonged Surgery	A072202	Low Impedance	20/02/2023	20/02/2023
								A090801	Signal Artifact/Noise	20/02/2023	20/02/2023
				CT3002	Implantable synthetic polymer	E2403	No Clinical Signs, Symptoms or Conditions	A072202	Low Impedance	20/02/2023	20/02/2023
								A090801	Signal Artifact/Noise	20/02/2023	20/02/2023
						F1905	Device Revision or Replacement	A072202	Low Impedance	20/02/2023	20/02/2023
								A090801	Signal Artifact/Noise	20/02/2023	20/02/2023
						F1908	Prolonged Surgery	A072202	Low Impedance	20/02/2023	20/02/2023
								A090801	Signal Artifact/Noise	20/02/2023	20/02/2023
				CT3006	Implantable inorganic material	E2403	No Clinical Signs, Symptoms or Conditions	A072202	Low Impedance	20/02/2023	20/02/2023
								A090801	Signal Artifact/Noise	20/02/2023	20/02/2023
						F1905	Device Revision or Replacement	A072202	Low Impedance	20/02/2023	20/02/2023
								A090801	Signal Artifact/Noise	20/02/2023	20/02/2023
						F1908	Prolonged Surgery	A072202	Low Impedance	20/02/2023	20/02/2023
								A090801	Signal Artifact/Noise	20/02/2023	20/02/2023
				CT314	Electrically powered	E2403	No Clinical Signs, Symptoms or Conditions	A072202	Low Impedance	20/02/2023	20/02/2023
								A090801	Signal Artifact/Noise	20/02/2023	20/02/2023
						F1905	Device Revision or Replacement	A072202	Low Impedance	20/02/2023	20/02/2023
								A090801	Signal Artifact/Noise	20/02/2023	20/02/2023
						F1908	Prolonged Surgery	A072202	Low Impedance	20/02/2023	20/02/2023
								A090801	Signal Artifact/Noise	20/02/2023	20/02/2023
				CT321	Long-term surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A072202	Low Impedance	20/02/2023	20/02/2023
								A090801	Signal Artifact/Noise	20/02/2023	20/02/2023
						F1905	Device Revision or Replacement	A072202	Low Impedance	20/02/2023	20/02/2023
								A090801	Signal Artifact/Noise	20/02/2023	20/02/2023
						F1908	Prolonged Surgery	A072202	Low Impedance	20/02/2023	20/02/2023
								A090801	Signal Artifact/Noise	20/02/2023	20/02/2023
				CT449	Generators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A072202	Low Impedance	20/02/2023	20/02/2023
								A090801	Signal Artifact/Noise	20/02/2023	20/02/2023
						F1905	Device Revision or Replacement	A072202	Low Impedance	20/02/2023	20/02/2023
								A090801	Signal Artifact/Noise	20/02/2023	20/02/2023
						F1908	Prolonged Surgery	A072202	Low Impedance	20/02/2023	20/02/2023
								A090801	Signal Artifact/Noise	20/02/2023	20/02/2023
				CT981	Single use	E2403	No Clinical Signs, Symptoms or Conditions	A072202	Low Impedance	20/02/2023	20/02/2023

									A090801	Signal Artifact/Noise	20/02/2023	20/02/2023
						F1905	Device Revision or Replacement		A072202	Low Impedance	20/02/2023	20/02/2023
									A090801	Signal Artifact/Noise	20/02/2023	20/02/2023
						F1908	Prolonged Surgery		A072202	Low Impedance	20/02/2023	20/02/2023
									A090801	Signal Artifact/Noise	20/02/2023	20/02/2023
				CT983	Surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions		A072202	Low Impedance	20/02/2023	20/02/2023
									A090801	Signal Artifact/Noise	20/02/2023	20/02/2023
						F1905	Device Revision or Replacement		A072202	Low Impedance	20/02/2023	20/02/2023
									A090801	Signal Artifact/Noise	20/02/2023	20/02/2023
						F1908	Prolonged Surgery		A072202	Low Impedance	20/02/2023	20/02/2023
									A090801	Signal Artifact/Noise	20/02/2023	20/02/2023
13285	2023	All other reportable incidents	36241	Pacemaker/defibrillator lead	CT1684	Cardiac pacemakers and associated devices	E0604	Cardiac Perforation		A0709	Device Sensing Problem		
										A0712	Pacing Problem		
										A1502	Positioning Problem		
							E0619	Pericardial Effusion		A0709	Device Sensing Problem		
										A0712	Pacing Problem		
										A1502	Positioning Problem		
							E0722	Hemothorax		A0709	Device Sensing Problem		
										A0712	Pacing Problem		
										A1502	Positioning Problem		
							F08	Hospitalization or Prolonged Hospitalization		A0709	Device Sensing Problem		
										A0712	Pacing Problem		
										A1502	Positioning Problem		
							F1203	Life Threatening Illness or Injury		A0709	Device Sensing Problem		
										A0712	Pacing Problem		
										A1502	Positioning Problem		
							F1904	Device Repositioning		A0709	Device Sensing Problem		
										A0712	Pacing Problem		
										A1502	Positioning Problem		
13286	2023	All other reportable incidents	47265	Dual-chamber implantable pacemaker, rate-responsive	CT141	Electrodes/Leads and associated devices	E2403	No Clinical Signs, Symptoms or Conditions		A24	No Apparent Device Problem		
							F1901	Additional Surgery		A24	No Apparent Device Problem		
					CT145	Cardiovascular devices	E2403	No Clinical Signs, Symptoms or Conditions		A24	No Apparent Device Problem		
							F1901	Additional Surgery		A24	No Apparent Device Problem		
					CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions		A24	No Apparent Device Problem		
							F1901	Additional Surgery		A24	No Apparent Device Problem		
					CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions		A24	No Apparent Device Problem		
							F1901	Additional Surgery		A24	No Apparent Device Problem		
					CT233	Surgical	E2403	No Clinical Signs, Symptoms or Conditions		A24	No Apparent Device Problem		
							F1901	Additional Surgery		A24	No Apparent Device Problem		
					CT2406	Implantable	E2403	No Clinical Signs, Symptoms or Conditions		A24	No Apparent Device Problem		
							F1901	Additional Surgery		A24	No Apparent Device Problem		
					CT2407	Active implantable	E2403	No Clinical Signs, Symptoms or Conditions		A24	No Apparent Device Problem		
							F1901	Additional Surgery		A24	No Apparent Device Problem		
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions		A24	No Apparent Device Problem		
							F1901	Additional Surgery		A24	No Apparent Device Problem		
					CT2727	Active implant devices/ancillaries	E2403	No Clinical Signs, Symptoms or Conditions		A24	No Apparent Device Problem		
							F1901	Additional Surgery		A24	No Apparent Device Problem		
					CT2750	Implantable cardiac electrical stimulators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions		A24	No Apparent Device Problem		
							F1901	Additional Surgery		A24	No Apparent Device Problem		
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions		A24	No Apparent Device Problem		
							F1901	Additional Surgery		A24	No Apparent Device Problem		
					CT2985	Active	E2403	No Clinical Signs, Symptoms or Conditions		A24	No Apparent Device Problem		
							F1901	Additional Surgery		A24	No Apparent Device Problem		
					CT3001	Implantable metal	E2403	No Clinical Signs, Symptoms or Conditions		A24	No Apparent Device Problem		
							F1901	Additional Surgery		A24	No Apparent Device Problem		
					CT3002	Implantable synthetic polymer	E2403	No Clinical Signs, Symptoms or Conditions		A24	No Apparent Device Problem		
							F1901	Additional Surgery		A24	No Apparent Device Problem		
					CT3006	Implantable inorganic material	E2403	No Clinical Signs, Symptoms or Conditions		A24	No Apparent Device Problem		
							F1901	Additional Surgery		A24	No Apparent Device Problem		
					CT314	Electrically powered	E2403	No Clinical Signs, Symptoms or Conditions		A24	No Apparent Device Problem		
							F1901	Additional Surgery		A24	No Apparent Device Problem		
					CT321	Long-term surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions		A24	No Apparent Device Problem		
							F1901	Additional Surgery		A24	No Apparent Device Problem		
					CT449	Generators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions		A24	No Apparent Device Problem		
							F1901	Additional Surgery		A24	No Apparent Device Problem		
					CT981	Single use	E2403	No Clinical Signs, Symptoms or Conditions		A24	No Apparent Device Problem		
							F1901	Additional Surgery		A24	No Apparent Device Problem		
					CT983	Surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions		A24	No Apparent Device Problem		
							F1901	Additional Surgery		A24	No Apparent Device Problem		
13287	2023	All other reportable incidents	35852	Single-chamber implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions		A070504	Premature Discharge of Battery	17/08/2023	17/08/2023
							F08	Hospitalization or Prolonged Hospitalization		A070504	Premature Discharge of Battery	17/08/2023	17/08/2023
							F1905	Device Revision or Replacement		A070504	Premature Discharge of Battery	17/08/2023	17/08/2023
13288	2023	All other reportable incidents	47265	Dual-chamber implantable pacemaker, rate-responsive	CT141	Electrodes/Leads and associated devices	E2403	No Clinical Signs, Symptoms or Conditions		A070504	Premature Discharge of Battery	24/05/2023	24/05/2023
							F08	Hospitalization or Prolonged Hospitalization		A070504	Premature Discharge of Battery	24/05/2023	24/05/2023
							F1905	Device Revision or Replacement		A070504	Premature Discharge of Battery	24/05/2023	24/05/2023
					CT145	Cardiovascular devices	E2403	No Clinical Signs, Symptoms or Conditions		A070504	Premature Discharge of Battery	24/05/2023	24/05/2023
							F08	Hospitalization or Prolonged Hospitalization		A070504	Premature Discharge of Battery	24/05/2023	24/05/2023
							F1905	Device Revision or Replacement		A070504	Premature Discharge of Battery	24/05/2023	24/05/2023
					CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions		A070504	Premature Discharge of Battery	24/05/2023	24/05/2023
							F08	Hospitalization or Prolonged Hospitalization		A070504	Premature Discharge of Battery	24/05/2023	24/05/2023
							F1905	Device Revision or Replacement		A070504	Premature Discharge of Battery	24/05/2023	24/05/2023
					CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions		A070504	Premature Discharge of Battery	24/05/2023	24/05/2023
							F08	Hospitalization or Prolonged Hospitalization		A070504	Premature Discharge of Battery	24/05/2023	24/05/2023
							F1905	Device Revision or Replacement		A070504	Premature Discharge of Battery	24/05/2023	24/05/2023
					CT233	Surgical	E2403	No Clinical Signs, Symptoms or Conditions		A070504	Premature Discharge of Battery	24/05/2023	24/05/2023
							F08	Hospitalization or Prolonged Hospitalization		A070504	Premature Discharge of Battery	24/05/2023	24/05/2023
							F1905	Device Revision or Replacement		A070504	Premature Discharge of Battery	24/05/2023	24/05/2023
					CT2406	Implantable	E2403	No Clinical Signs, Symptoms or Conditions		A070504	Premature Discharge of Battery	24/05/2023	24/05/2023

							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	24/05/2023	24/05/2023
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	24/05/2023	24/05/2023
					CT2407	Active implantable	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	24/05/2023	24/05/2023
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	24/05/2023	24/05/2023
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	24/05/2023	24/05/2023
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	24/05/2023	24/05/2023
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	24/05/2023	24/05/2023
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	24/05/2023	24/05/2023
					CT2727	Active implant devices/ancillaries	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	24/05/2023	24/05/2023
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	24/05/2023	24/05/2023
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	24/05/2023	24/05/2023
					CT2750	Implantable cardiac electrical stimulators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	24/05/2023	24/05/2023
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	24/05/2023	24/05/2023
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	24/05/2023	24/05/2023
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	24/05/2023	24/05/2023
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	24/05/2023	24/05/2023
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	24/05/2023	24/05/2023
					CT2985	Active	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	24/05/2023	24/05/2023
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	24/05/2023	24/05/2023
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	24/05/2023	24/05/2023
					CT3001	Implantable metal	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	24/05/2023	24/05/2023
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	24/05/2023	24/05/2023
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	24/05/2023	24/05/2023
					CT3002	Implantable synthetic polymer	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	24/05/2023	24/05/2023
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	24/05/2023	24/05/2023
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	24/05/2023	24/05/2023
					CT3006	Implantable inorganic material	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	24/05/2023	24/05/2023
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	24/05/2023	24/05/2023
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	24/05/2023	24/05/2023
					CT314	Electrically powered	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	24/05/2023	24/05/2023
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	24/05/2023	24/05/2023
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	24/05/2023	24/05/2023
					CT321	Long-term surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	24/05/2023	24/05/2023
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	24/05/2023	24/05/2023
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	24/05/2023	24/05/2023
					CT449	Generators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	24/05/2023	24/05/2023
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	24/05/2023	24/05/2023
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	24/05/2023	24/05/2023
					CT981	Single use	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	24/05/2023	24/05/2023
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	24/05/2023	24/05/2023
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	24/05/2023	24/05/2023
					CT983	Surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	24/05/2023	24/05/2023
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	24/05/2023	24/05/2023
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	24/05/2023	24/05/2023
13289	2023	All other reportable incidents	47267	Single-chamber implantable pacemaker, rate-responsive	CT141	Electrodes/Leads and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070101	Failure to Capture		
							F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A070101	Failure to Capture		
					CT145	Cardiovascular devices	E2403	No Clinical Signs, Symptoms or Conditions	A070101	Failure to Capture		
							F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A070101	Failure to Capture		
					CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070101	Failure to Capture		
							F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A070101	Failure to Capture		
					CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A070101	Failure to Capture		
							F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A070101	Failure to Capture		
					CT233	Surgical	E2403	No Clinical Signs, Symptoms or Conditions	A070101	Failure to Capture		
							F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A070101	Failure to Capture		
					CT2406	Implantable	E2403	No Clinical Signs, Symptoms or Conditions	A070101	Failure to Capture		
							F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A070101	Failure to Capture		
					CT2407	Active implantable	E2403	No Clinical Signs, Symptoms or Conditions	A070101	Failure to Capture		
							F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A070101	Failure to Capture		
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A070101	Failure to Capture		
							F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A070101	Failure to Capture		
					CT2727	Active implant devices/ancillaries	E2403	No Clinical Signs, Symptoms or Conditions	A070101	Failure to Capture		
							F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A070101	Failure to Capture		
					CT2750	Implantable cardiac electrical stimulators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070101	Failure to Capture		
							F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A070101	Failure to Capture		
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A070101	Failure to Capture		
							F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A070101	Failure to Capture		
					CT2985	Active	E2403	No Clinical Signs, Symptoms or Conditions	A070101	Failure to Capture		
							F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A070101	Failure to Capture		
					CT3001	Implantable metal	E2403	No Clinical Signs, Symptoms or Conditions	A070101	Failure to Capture		
							F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A070101	Failure to Capture		
					CT3002	Implantable synthetic polymer	E2403	No Clinical Signs, Symptoms or Conditions	A070101	Failure to Capture		
							F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A070101	Failure to Capture		
					CT3006	Implantable inorganic material	E2403	No Clinical Signs, Symptoms or Conditions	A070101	Failure to Capture		
							F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A070101	Failure to Capture		
					CT314	Electrically powered	E2403	No Clinical Signs, Symptoms or Conditions	A070101	Failure to Capture		
							F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A070101	Failure to Capture		
					CT321	Long-term surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A070101	Failure to Capture		
							F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A070101	Failure to Capture		
					CT449	Generators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070101	Failure to Capture		
							F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A070101	Failure to Capture		
					CT981	Single use	E2403	No Clinical Signs, Symptoms or Conditions	A070101	Failure to Capture		
							F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A070101	Failure to Capture		
					CT983	Surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A070101	Failure to Capture		
							F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A070101	Failure to Capture		
13290	2023	All other reportable incidents	17882	Physiologic-monitoring defibrillation system	CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A0719	Unexpected Shutdown		
							F26	No Health Consequences or Impact	A0719	Unexpected Shutdown		
13291	2023	All other reportable incidents	47265	Dual-chamber implantable pacemaker, rate-responsive	CT141	Electrodes/Leads and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		

					CT145	Cardiovascular devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT233	Surgical	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT2406	Implantable	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT2407	Active implantable	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT2727	Active implant devices/ancillaries	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT2750	Implantable cardiac electrical stimulators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT2985	Active	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT3001	Implantable metal	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT3002	Implantable synthetic polymer	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT3006	Implantable inorganic material	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT314	Electrically powered	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT321	Long-term surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT449	Generators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT981	Single use	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT983	Surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
13292	2023	All other reportable incidents	17882	Physiologic-monitoring defibrillation system	CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A0719	Unexpected Shutdown		
							F27	Problem Identified During Non-Clinical Procedure	A0719	Unexpected Shutdown		
13293	2023	All other reportable incidents	47205	Cardiac pulse generator programmer	CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A150105	Key or Button Unresponsive/not Working		
							F26	No Health Consequences or Impact	A150105	Key or Button Unresponsive/not Working		
13294	2023	All other reportable incidents	35852	Single-chamber implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	16/06/2025	16/06/2025
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	16/06/2025	16/06/2025
							F1903	Device Explantation	A070504	Premature Discharge of Battery	16/06/2025	16/06/2025
13295	2023		47265	Dual-chamber implantable pacemaker, rate-responsive	CT141	Electrodes/Leads and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	23/05/2023	23/05/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	23/05/2023	23/05/2023
					CT145	Cardiovascular devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	23/05/2023	23/05/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	23/05/2023	23/05/2023
					CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	23/05/2023	23/05/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	23/05/2023	23/05/2023
					CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	23/05/2023	23/05/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	23/05/2023	23/05/2023
					CT233	Surgical	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	23/05/2023	23/05/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	23/05/2023	23/05/2023
					CT2406	Implantable	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	23/05/2023	23/05/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	23/05/2023	23/05/2023
					CT2407	Active implantable	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	23/05/2023	23/05/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	23/05/2023	23/05/2023
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	23/05/2023	23/05/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	23/05/2023	23/05/2023
					CT2727	Active implant devices/ancillaries	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	23/05/2023	23/05/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	23/05/2023	23/05/2023
					CT2750	Implantable cardiac electrical stimulators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	23/05/2023	23/05/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	23/05/2023	23/05/2023
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	23/05/2023	23/05/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	23/05/2023	23/05/2023
					CT2985	Active	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	23/05/2023	23/05/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	23/05/2023	23/05/2023
					CT3001	Implantable metal	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	23/05/2023	23/05/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	23/05/2023	23/05/2023
					CT3002	Implantable synthetic polymer	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	23/05/2023	23/05/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	23/05/2023	23/05/2023
					CT3006	Implantable inorganic material	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	23/05/2023	23/05/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	23/05/2023	23/05/2023
					CT314	Electrically powered	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	23/05/2023	23/05/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	23/05/2023	23/05/2023
					CT321	Long-term surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	23/05/2023	23/05/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	23/05/2023	23/05/2023
					CT449	Generators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	23/05/2023	23/05/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	23/05/2023	23/05/2023
					CT981	Single use	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	23/05/2023	23/05/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	23/05/2023	23/05/2023
					CT983	Surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	23/05/2023	23/05/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	23/05/2023	23/05/2023
13296	2023	All other reportable incidents	17882	Physiologic-monitoring defibrillation system	CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A071301	Failure to Deliver Shock/Stimulation		
							F26	No Health Consequences or Impact	A071301	Failure to Deliver Shock/Stimulation		
13297	2023	All other reportable incidents	17882	Physiologic-monitoring defibrillation system	CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070803	Failure to Power Up		

							F27	Problem Identified During Non-Clinical Procedure	A070803	Failure to Power Up		
13298	2023		47265	Dual-chamber implantable pacemaker, rate-responsive	CT141	Electrodes/Leads and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	04/05/2023	04/05/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	04/05/2023	04/05/2023
					CT145	Cardiovascular devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	04/05/2023	04/05/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	04/05/2023	04/05/2023
					CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	04/05/2023	04/05/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	04/05/2023	04/05/2023
					CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	04/05/2023	04/05/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	04/05/2023	04/05/2023
					CT233	Surgical	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	04/05/2023	04/05/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	04/05/2023	04/05/2023
					CT2406	Implantable	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	04/05/2023	04/05/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	04/05/2023	04/05/2023
					CT2407	Active implantable	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	04/05/2023	04/05/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	04/05/2023	04/05/2023
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	04/05/2023	04/05/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	04/05/2023	04/05/2023
					CT2727	Active implant devices/ancillaries	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	04/05/2023	04/05/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	04/05/2023	04/05/2023
					CT2750	Implantable cardiac electrical stimulators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	04/05/2023	04/05/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	04/05/2023	04/05/2023
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	04/05/2023	04/05/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	04/05/2023	04/05/2023
					CT2985	Active	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	04/05/2023	04/05/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	04/05/2023	04/05/2023
					CT3001	Implantable metal	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	04/05/2023	04/05/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	04/05/2023	04/05/2023
					CT3002	Implantable synthetic polymer	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	04/05/2023	04/05/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	04/05/2023	04/05/2023
					CT3006	Implantable inorganic material	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	04/05/2023	04/05/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	04/05/2023	04/05/2023
					CT314	Electrically powered	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	04/05/2023	04/05/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	04/05/2023	04/05/2023
					CT321	Long-term surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	04/05/2023	04/05/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	04/05/2023	04/05/2023
					CT449	Generators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	04/05/2023	04/05/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	04/05/2023	04/05/2023
					CT981	Single use	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	04/05/2023	04/05/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	04/05/2023	04/05/2023
					CT983	Surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	04/05/2023	04/05/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	04/05/2023	04/05/2023
13299	2023		47265	Dual-chamber implantable pacemaker, rate-responsive	CT141	Electrodes/Leads and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	04/05/2023	04/05/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	04/05/2023	04/05/2023
					CT145	Cardiovascular devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	04/05/2023	04/05/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	04/05/2023	04/05/2023
					CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	04/05/2023	04/05/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	04/05/2023	04/05/2023
					CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	04/05/2023	04/05/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	04/05/2023	04/05/2023
					CT233	Surgical	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	04/05/2023	04/05/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	04/05/2023	04/05/2023
					CT2406	Implantable	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	04/05/2023	04/05/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	04/05/2023	04/05/2023
					CT2407	Active implantable	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	04/05/2023	04/05/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	04/05/2023	04/05/2023
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	04/05/2023	04/05/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	04/05/2023	04/05/2023
					CT2727	Active implant devices/ancillaries	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	04/05/2023	04/05/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	04/05/2023	04/05/2023
					CT2750	Implantable cardiac electrical stimulators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	04/05/2023	04/05/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	04/05/2023	04/05/2023
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	04/05/2023	04/05/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	04/05/2023	04/05/2023
					CT2985	Active	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	04/05/2023	04/05/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	04/05/2023	04/05/2023
					CT3001	Implantable metal	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	04/05/2023	04/05/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	04/05/2023	04/05/2023
					CT3002	Implantable synthetic polymer	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	04/05/2023	04/05/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	04/05/2023	04/05/2023
					CT3006	Implantable inorganic material	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	04/05/2023	04/05/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	04/05/2023	04/05/2023
					CT314	Electrically powered	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	04/05/2023	04/05/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	04/05/2023	04/05/2023
					CT321	Long-term surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	04/05/2023	04/05/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	04/05/2023	04/05/2023
					CT449	Generators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	04/05/2023	04/05/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	04/05/2023	04/05/2023
					CT981	Single use	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	04/05/2023	04/05/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	04/05/2023	04/05/2023
					CT983	Surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	04/05/2023	04/05/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	04/05/2023	04/05/2023
13300	2023	All other reportable incidents	35852	Single-chamber implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
									A0908	Incorrect, Inadequate or Imprecise Result or Readings		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
									A0908	Incorrect, Inadequate or Imprecise Result or Readings		
13301	2023	All other reportable incidents	17882	Physiologic-monitoring defibrillation system	CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070803	Failure to Power Up		

							F27	Problem Identified During Non-Clinical Procedure	A070803	Failure to Power Up		
13302	2023	All other reportable incidents	47205	Cardiac pulse generator programmer	CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A150105	Key or Button Unresponsive/not Working		
							F26	No Health Consequences or Impact	A150105	Key or Button Unresponsive/not Working		
13303	2023	All other reportable incidents	37265	Dual-chamber implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A070909	Oversensing		
							F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A070909	Oversensing		
13304	2023	All other reportable incidents	36241	Pacemaker/defibrillator lead	CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A04	Material Integrity Problem	30/05/2023	30/05/2023
									A05	Mechanical Problem	30/05/2023	30/05/2023
							F1908	Prolonged Surgery	A04	Material Integrity Problem	30/05/2023	30/05/2023
									A05	Mechanical Problem	30/05/2023	30/05/2023
13305	2023	All other reportable incidents	36241	Pacemaker/defibrillator lead	CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A072201	High Impedance		
							F26	No Health Consequences or Impact	A072201	High Impedance		
13306	2023	All other reportable incidents	47263	Cardiac resynchronization therapy implantable pacemaker	CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070909	Oversensing		
									A072201	High Impedance		
									A090801	Signal Artifact/Noise		
							F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A070909	Oversensing		
									A072201	High Impedance		
									A090801	Signal Artifact/Noise		
					CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A070909	Oversensing		
									A072201	High Impedance		
									A090801	Signal Artifact/Noise		
							F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A070909	Oversensing		
									A072201	High Impedance		
									A090801	Signal Artifact/Noise		
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A070909	Oversensing		
									A072201	High Impedance		
									A090801	Signal Artifact/Noise		
							F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A070909	Oversensing		
									A072201	High Impedance		
									A090801	Signal Artifact/Noise		
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A070909	Oversensing		
									A072201	High Impedance		
									A090801	Signal Artifact/Noise		
							F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A070909	Oversensing		
									A072201	High Impedance		
									A090801	Signal Artifact/Noise		
13307	2023	All other reportable incidents	36241	Pacemaker/defibrillator lead	CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A040101	Fracture		
									A070101	Failure to Capture		
									A070909	Oversensing		
									A0712	Pacing Problem		
									A072201	High Impedance		
									A090801	Signal Artifact/Noise		
							F08	Hospitalization or Prolonged Hospitalization	A040101	Fracture		
									A070101	Failure to Capture		
									A070909	Oversensing		
									A0712	Pacing Problem		
									A072201	High Impedance		
									A090801	Signal Artifact/Noise		
							F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A040101	Fracture		
									A070101	Failure to Capture		
									A070909	Oversensing		
									A0712	Pacing Problem		
									A072201	High Impedance		
									A090801	Signal Artifact/Noise		
							F1905	Device Revision or Replacement	A040101	Fracture		
									A070101	Failure to Capture		
									A070909	Oversensing		
									A0712	Pacing Problem		
									A072201	High Impedance		
									A090801	Signal Artifact/Noise		
							F2203	Imaging Required	A040101	Fracture		
									A070101	Failure to Capture		
									A070909	Oversensing		
									A0712	Pacing Problem		
									A072201	High Impedance		
									A090801	Signal Artifact/Noise		
13308	2023	All other reportable incidents	35223	Endocardial/interventricular septal pacing lead	CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070909	Oversensing		
							F26	No Health Consequences or Impact	A070909	Oversensing		
13309	2023	All other reportable incidents	47205	Cardiac pulse generator programmer	CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A1102	Application Program Problem		
									A150105	Key or Button Unresponsive/not Working		
							F26	No Health Consequences or Impact	A1102	Application Program Problem		
									A150105	Key or Button Unresponsive/not Working		
13310	2023		47265	Dual-chamber implantable pacemaker, rate-responsive	CT141	Electrodes/Leads and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem		
							F1901	Additional Surgery	A24	No Apparent Device Problem		
					CT145	Cardiovascular devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem		
							F1901	Additional Surgery	A24	No Apparent Device Problem		
					CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem		
							F1901	Additional Surgery	A24	No Apparent Device Problem		
					CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem		
							F1901	Additional Surgery	A24	No Apparent Device Problem		
					CT233	Surgical	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem		
							F1901	Additional Surgery	A24	No Apparent Device Problem		
					CT2406	Implantable	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem		
							F1901	Additional Surgery	A24	No Apparent Device Problem		
					CT2407	Active implantable	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem		
							F1901	Additional Surgery	A24	No Apparent Device Problem		
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem		
							F1901	Additional Surgery	A24	No Apparent Device Problem		
					CT2727	Active implant devices/ancillaries	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem		

							F1901	Additional Surgery	A24	No Apparent Device Problem		
					CT2750	Implantable cardiac electrical stimulators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem		
							F1901	Additional Surgery	A24	No Apparent Device Problem		
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem		
							F1901	Additional Surgery	A24	No Apparent Device Problem		
					CT2985	Active	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem		
							F1901	Additional Surgery	A24	No Apparent Device Problem		
					CT3001	Implantable metal	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem		
							F1901	Additional Surgery	A24	No Apparent Device Problem		
					CT3002	Implantable synthetic polymer	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem		
							F1901	Additional Surgery	A24	No Apparent Device Problem		
					CT3006	Implantable inorganic material	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem		
							F1901	Additional Surgery	A24	No Apparent Device Problem		
					CT314	Electrically powered	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem		
							F1901	Additional Surgery	A24	No Apparent Device Problem		
					CT321	Long-term surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem		
							F1901	Additional Surgery	A24	No Apparent Device Problem		
					CT449	Generators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem		
							F1901	Additional Surgery	A24	No Apparent Device Problem		
					CT981	Single use	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem		
							F1901	Additional Surgery	A24	No Apparent Device Problem		
					CT983	Surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem		
							F1901	Additional Surgery	A24	No Apparent Device Problem		
13311	2023		47265	Dual-chamber implantable pacemaker, rate-responsive	CT141	Electrodes/Leads and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	31/05/2023	31/05/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	31/05/2023	31/05/2023
					CT145	Cardiovascular devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	31/05/2023	31/05/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	31/05/2023	31/05/2023
					CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	31/05/2023	31/05/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	31/05/2023	31/05/2023
					CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	31/05/2023	31/05/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	31/05/2023	31/05/2023
					CT233	Surgical	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	31/05/2023	31/05/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	31/05/2023	31/05/2023
					CT2406	Implantable	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	31/05/2023	31/05/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	31/05/2023	31/05/2023
					CT2407	Active implantable	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	31/05/2023	31/05/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	31/05/2023	31/05/2023
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	31/05/2023	31/05/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	31/05/2023	31/05/2023
					CT2727	Active implant devices/ancillaries	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	31/05/2023	31/05/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	31/05/2023	31/05/2023
					CT2750	Implantable cardiac electrical stimulators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	31/05/2023	31/05/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	31/05/2023	31/05/2023
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	31/05/2023	31/05/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	31/05/2023	31/05/2023
					CT2985	Active	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	31/05/2023	31/05/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	31/05/2023	31/05/2023
					CT3001	Implantable metal	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	31/05/2023	31/05/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	31/05/2023	31/05/2023
					CT3002	Implantable synthetic polymer	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	31/05/2023	31/05/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	31/05/2023	31/05/2023
					CT3006	Implantable inorganic material	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	31/05/2023	31/05/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	31/05/2023	31/05/2023
					CT314	Electrically powered	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	31/05/2023	31/05/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	31/05/2023	31/05/2023
					CT321	Long-term surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	31/05/2023	31/05/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	31/05/2023	31/05/2023
					CT449	Generators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	31/05/2023	31/05/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	31/05/2023	31/05/2023
					CT981	Single use	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	31/05/2023	31/05/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	31/05/2023	31/05/2023
					CT983	Surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	31/05/2023	31/05/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	31/05/2023	31/05/2023
13312	2023	All other reportable incidents	47263	Cardiac resynchronization therapy implantable pacemaker	CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	20/12/2022	20/12/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	20/12/2022	20/12/2022
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	20/12/2022	20/12/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	20/12/2022	20/12/2022
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	20/12/2022	20/12/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	20/12/2022	20/12/2022
13313	2023	All other reportable incidents	47263	Cardiac resynchronization therapy implantable pacemaker	CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	05/01/2023	05/01/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	05/01/2023	05/01/2023
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	05/01/2023	05/01/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	05/01/2023	05/01/2023
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	05/01/2023	05/01/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	05/01/2023	05/01/2023
13314	2023	All other reportable incidents	47263	Cardiac resynchronization therapy implantable pacemaker	CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	05/01/2023	05/01/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	05/01/2023	05/01/2023
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	05/01/2023	05/01/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	05/01/2023	05/01/2023
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	05/01/2023	05/01/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	05/01/2023	05/01/2023
13315	2023	All other reportable incidents	47265	Dual-chamber implantable pacemaker, rate-responsive	CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A0908	Incorrect, Inadequate or Imprecise Result or Readings	05/06/2023	05/06/2023
							F08	Hospitalization or Prolonged Hospitalization	A0908	Incorrect, Inadequate or Imprecise Result or Readings	05/06/2023	05/06/2023
							F1905	Device Revision or Replacement	A0908	Incorrect, Inadequate or Imprecise Result or Readings	05/06/2023	05/06/2023
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A0908	Incorrect, Inadequate or Imprecise Result or Readings	05/06/2023	05/06/2023

							F08	Hospitalization or Prolonged Hospitalization	A0908	Incorrect, Inadequate or Imprecise Result or Readings	05/06/2023	05/06/2023
							F1905	Device Revision or Replacement	A0908	Incorrect, Inadequate or Imprecise Result or Readings	05/06/2023	05/06/2023
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A0908	Incorrect, Inadequate or Imprecise Result or Readings	05/06/2023	05/06/2023
							F08	Hospitalization or Prolonged Hospitalization	A0908	Incorrect, Inadequate or Imprecise Result or Readings	05/06/2023	05/06/2023
							F1905	Device Revision or Replacement	A0908	Incorrect, Inadequate or Imprecise Result or Readings	05/06/2023	05/06/2023
13316	2023	All other reportable incidents	47265	Dual-chamber implantable pacemaker, rate-responsive	CT141	Electrodes/Leads and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem		
							F1901	Additional Surgery	A24	No Apparent Device Problem		
					CT145	Cardiovascular devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem		
							F1901	Additional Surgery	A24	No Apparent Device Problem		
					CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem		
							F1901	Additional Surgery	A24	No Apparent Device Problem		
					CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem		
							F1901	Additional Surgery	A24	No Apparent Device Problem		
					CT233	Surgical	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem		
							F1901	Additional Surgery	A24	No Apparent Device Problem		
					CT2406	Implantable	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem		
							F1901	Additional Surgery	A24	No Apparent Device Problem		
					CT2407	Active implantable	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem		
							F1901	Additional Surgery	A24	No Apparent Device Problem		
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem		
							F1901	Additional Surgery	A24	No Apparent Device Problem		
					CT2727	Active implant devices/ancillaries	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem		
							F1901	Additional Surgery	A24	No Apparent Device Problem		
					CT2750	Implantable cardiac electrical stimulators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem		
							F1901	Additional Surgery	A24	No Apparent Device Problem		
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem		
							F1901	Additional Surgery	A24	No Apparent Device Problem		
					CT2985	Active	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem		
							F1901	Additional Surgery	A24	No Apparent Device Problem		
					CT3001	Implantable metal	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem		
							F1901	Additional Surgery	A24	No Apparent Device Problem		
					CT3002	Implantable synthetic polymer	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem		
							F1901	Additional Surgery	A24	No Apparent Device Problem		
					CT3006	Implantable inorganic material	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem		
							F1901	Additional Surgery	A24	No Apparent Device Problem		
					CT314	Electrically powered	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem		
							F1901	Additional Surgery	A24	No Apparent Device Problem		
					CT321	Long-term surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem		
							F1901	Additional Surgery	A24	No Apparent Device Problem		
					CT449	Generators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem		
							F1901	Additional Surgery	A24	No Apparent Device Problem		
					CT981	Single use	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem		
							F1901	Additional Surgery	A24	No Apparent Device Problem		
					CT983	Surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem		
							F1901	Additional Surgery	A24	No Apparent Device Problem		
13317	2023	All other reportable incidents	35223	Endocardial/interventricular septal pacing lead	CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A05	Mechanical Problem		
									A0701	Capturing Problem		
									A070909	Oversensing		
									A0722	Impedance Problem		
							F26	No Health Consequences or Impact	A05	Mechanical Problem		
									A0701	Capturing Problem		
									A070909	Oversensing		
									A0722	Impedance Problem		
13318	2023	All other reportable incidents	47265	Dual-chamber implantable pacemaker, rate-responsive	CT141	Electrodes/Leads and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem		
							F26	No Health Consequences or Impact	A24	No Apparent Device Problem		
					CT145	Cardiovascular devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem		
							F26	No Health Consequences or Impact	A24	No Apparent Device Problem		
					CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem		
							F26	No Health Consequences or Impact	A24	No Apparent Device Problem		
					CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem		
							F26	No Health Consequences or Impact	A24	No Apparent Device Problem		
					CT233	Surgical	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem		
							F26	No Health Consequences or Impact	A24	No Apparent Device Problem		
					CT2406	Implantable	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem		
							F26	No Health Consequences or Impact	A24	No Apparent Device Problem		
					CT2407	Active implantable	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem		
							F26	No Health Consequences or Impact	A24	No Apparent Device Problem		
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem		
							F26	No Health Consequences or Impact	A24	No Apparent Device Problem		
					CT2727	Active implant devices/ancillaries	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem		
							F26	No Health Consequences or Impact	A24	No Apparent Device Problem		
					CT2750	Implantable cardiac electrical stimulators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem		
							F26	No Health Consequences or Impact	A24	No Apparent Device Problem		
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem		
							F26	No Health Consequences or Impact	A24	No Apparent Device Problem		
					CT2985	Active	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem		
							F26	No Health Consequences or Impact	A24	No Apparent Device Problem		
					CT3001	Implantable metal	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem		
							F26	No Health Consequences or Impact	A24	No Apparent Device Problem		
					CT3002	Implantable synthetic polymer	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem		
							F26	No Health Consequences or Impact	A24	No Apparent Device Problem		
					CT3006	Implantable inorganic material	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem		
							F26	No Health Consequences or Impact	A24	No Apparent Device Problem		
					CT314	Electrically powered	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem		

							F26	No Health Consequences or Impact	A24	No Apparent Device Problem			
					CT321	Long-term surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem			
							F26	No Health Consequences or Impact	A24	No Apparent Device Problem			
					CT449	Generators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem			
							F26	No Health Consequences or Impact	A24	No Apparent Device Problem			
					CT981	Single use	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem			
							F26	No Health Consequences or Impact	A24	No Apparent Device Problem			
					CT983	Surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem			
							F26	No Health Consequences or Impact	A24	No Apparent Device Problem			
13319	2023		47265	Dual-chamber implantable pacemaker, rate-responsive	CT141	Electrodes/Leads and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	30/05/2023	30/05/2023	
							F1901	Additional Surgery	A24	No Apparent Device Problem	30/05/2023	30/05/2023	
					CT145	Cardiovascular devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	30/05/2023	30/05/2023	
							F1901	Additional Surgery	A24	No Apparent Device Problem	30/05/2023	30/05/2023	
					CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	30/05/2023	30/05/2023	
							F1901	Additional Surgery	A24	No Apparent Device Problem	30/05/2023	30/05/2023	
					CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	30/05/2023	30/05/2023	
							F1901	Additional Surgery	A24	No Apparent Device Problem	30/05/2023	30/05/2023	
					CT233	Surgical	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	30/05/2023	30/05/2023	
							F1901	Additional Surgery	A24	No Apparent Device Problem	30/05/2023	30/05/2023	
					CT2406	Implantable	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	30/05/2023	30/05/2023	
							F1901	Additional Surgery	A24	No Apparent Device Problem	30/05/2023	30/05/2023	
					CT2407	Active implantable	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	30/05/2023	30/05/2023	
							F1901	Additional Surgery	A24	No Apparent Device Problem	30/05/2023	30/05/2023	
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	30/05/2023	30/05/2023	
							F1901	Additional Surgery	A24	No Apparent Device Problem	30/05/2023	30/05/2023	
					CT2727	Active implant devices/ancillaries	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	30/05/2023	30/05/2023	
							F1901	Additional Surgery	A24	No Apparent Device Problem	30/05/2023	30/05/2023	
					CT2750	Implantable cardiac electrical stimulators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	30/05/2023	30/05/2023	
							F1901	Additional Surgery	A24	No Apparent Device Problem	30/05/2023	30/05/2023	
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	30/05/2023	30/05/2023	
							F1901	Additional Surgery	A24	No Apparent Device Problem	30/05/2023	30/05/2023	
					CT2985	Active	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	30/05/2023	30/05/2023	
							F1901	Additional Surgery	A24	No Apparent Device Problem	30/05/2023	30/05/2023	
					CT3001	Implantable metal	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	30/05/2023	30/05/2023	
							F1901	Additional Surgery	A24	No Apparent Device Problem	30/05/2023	30/05/2023	
					CT3002	Implantable synthetic polymer	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	30/05/2023	30/05/2023	
							F1901	Additional Surgery	A24	No Apparent Device Problem	30/05/2023	30/05/2023	
					CT3006	Implantable inorganic material	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	30/05/2023	30/05/2023	
							F1901	Additional Surgery	A24	No Apparent Device Problem	30/05/2023	30/05/2023	
					CT314	Electrically powered	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	30/05/2023	30/05/2023	
							F1901	Additional Surgery	A24	No Apparent Device Problem	30/05/2023	30/05/2023	
					CT321	Long-term surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	30/05/2023	30/05/2023	
							F1901	Additional Surgery	A24	No Apparent Device Problem	30/05/2023	30/05/2023	
					CT449	Generators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	30/05/2023	30/05/2023	
							F1901	Additional Surgery	A24	No Apparent Device Problem	30/05/2023	30/05/2023	
					CT981	Single use	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	30/05/2023	30/05/2023	
							F1901	Additional Surgery	A24	No Apparent Device Problem	30/05/2023	30/05/2023	
					CT983	Surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	30/05/2023	30/05/2023	
							F1901	Additional Surgery	A24	No Apparent Device Problem	30/05/2023	30/05/2023	
13320	2023	All other reportable incidents	35852	Single-chamber implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	31/05/2023	31/05/2023	
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	31/05/2023	31/05/2023	
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	31/05/2023	31/05/2023	
13321	2023	All other reportable incidents	47265	Dual-chamber implantable pacemaker, rate-responsive	CT141	Electrodes/Leads and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem			
									A070504	Premature Discharge of Battery			
							F26	No Health Consequences or Impact	A0705	Battery Problem			
									A070504	Premature Discharge of Battery			
					CT145	Cardiovascular devices	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem			
									A070504	Premature Discharge of Battery			
							F26	No Health Consequences or Impact	A0705	Battery Problem			
									A070504	Premature Discharge of Battery			
					CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem			
									A070504	Premature Discharge of Battery			
							F26	No Health Consequences or Impact	A0705	Battery Problem			
									A070504	Premature Discharge of Battery			
					CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem			
									A070504	Premature Discharge of Battery			
							F26	No Health Consequences or Impact	A0705	Battery Problem			
									A070504	Premature Discharge of Battery			
					CT233	Surgical	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem			
									A070504	Premature Discharge of Battery			
							F26	No Health Consequences or Impact	A0705	Battery Problem			
									A070504	Premature Discharge of Battery			
					CT2406	Implantable	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem			
									A070504	Premature Discharge of Battery			
							F26	No Health Consequences or Impact	A0705	Battery Problem			
									A070504	Premature Discharge of Battery			
					CT2407	Active implantable	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem			
									A070504	Premature Discharge of Battery			
							F26	No Health Consequences or Impact	A0705	Battery Problem			
									A070504	Premature Discharge of Battery			
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem			
									A070504	Premature Discharge of Battery			
							F26	No Health Consequences or Impact	A0705	Battery Problem			
									A070504	Premature Discharge of Battery			
					CT2727	Active implant devices/ancillaries	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem			
									A070504	Premature Discharge of Battery			
							F26	No Health Consequences or Impact	A0705	Battery Problem			
									A070504	Premature Discharge of Battery			

					CT2750	Implantable cardiac electrical stimulators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
					CT2985	Active	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
					CT3001	Implantable metal	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
					CT3002	Implantable synthetic polymer	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
					CT3006	Implantable inorganic material	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
					CT314	Electrically powered	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
					CT321	Long-term surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
					CT449	Generators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
					CT981	Single use	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
					CT983	Surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
13322	2023	All other reportable incidents	37265	Dual-chamber implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
							F24	Insufficient Health Impact Information	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
13323	2023		47265	Dual-chamber implantable pacemaker, rate-responsive	CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	23/03/2023	23/03/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	23/03/2023	23/03/2023
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	23/03/2023	23/03/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	23/03/2023	23/03/2023
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	23/03/2023	23/03/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	23/03/2023	23/03/2023
13324	2023	All other reportable incidents	17882	Physiologic-monitoring defibrillation system	CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A071204	No Pacing		
							F27	Problem Identified During Non-Clinical Procedure	A071204	No Pacing		
13325	2023	All other reportable incidents	37265	Dual-chamber implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A070909	Oversensing		
							F26	No Health Consequences or Impact	A070909	Oversensing		
13326	2023	All other reportable incidents	63188	Left-ventricular wireless pacing system sensor/transmitter	CT145	Cardiovascular devices	F24	Insufficient Health Impact Information	A070801	Complete Loss of Power		
					CT1577	Wireless monitoring system transmitters	F24	Insufficient Health Impact Information	A070801	Complete Loss of Power		
					CT1684	Cardiac pacemakers and associated devices	F24	Insufficient Health Impact Information	A070801	Complete Loss of Power		
					CT2727	Active implant devices/ancillaries	F24	Insufficient Health Impact Information	A070801	Complete Loss of Power		
					CT2750	Implantable cardiac electrical stimulators and associated devices	F24	Insufficient Health Impact Information	A070801	Complete Loss of Power		
					CT2885	Physiological monitoring/measuring system transmitters/receivers/transceivers	F24	Insufficient Health Impact Information	A070801	Complete Loss of Power		
13327	2023	All other reportable incidents	35852	Single-chamber implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A0508	Noise, Audible		
									A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A0508	Noise, Audible		
									A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
13328	2023	All other reportable incidents	47265	Dual-chamber implantable pacemaker, rate-responsive	CT141	Electrodes/Leads and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A0908	Incorrect, Inadequate or Imprecise Result or Readings		
							F26	No Health Consequences or Impact	A0908	Incorrect, Inadequate or Imprecise Result or Readings		
					CT145	Cardiovascular devices	E2403	No Clinical Signs, Symptoms or Conditions	A0908	Incorrect, Inadequate or Imprecise Result or Readings		
							F26	No Health Consequences or Impact	A0908	Incorrect, Inadequate or Imprecise Result or Readings		
					CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A0908	Incorrect, Inadequate or Imprecise Result or Readings		
							F26	No Health Consequences or Impact	A0908	Incorrect, Inadequate or Imprecise Result or Readings		
					CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A0908	Incorrect, Inadequate or Imprecise Result or Readings		
							F26	No Health Consequences or Impact	A0908	Incorrect, Inadequate or Imprecise Result or Readings		
					CT233	Surgical	E2403	No Clinical Signs, Symptoms or Conditions	A0908	Incorrect, Inadequate or Imprecise Result or Readings		

							F26	No Health Consequences or Impact	A0908	Incorrect, Inadequate or Imprecise Result or Readings			
					CT2406	Implantable	E2403	No Clinical Signs, Symptoms or Conditions	A0908	Incorrect, Inadequate or Imprecise Result or Readings			
							F26	No Health Consequences or Impact	A0908	Incorrect, Inadequate or Imprecise Result or Readings			
					CT2407	Active implantable	E2403	No Clinical Signs, Symptoms or Conditions	A0908	Incorrect, Inadequate or Imprecise Result or Readings			
							F26	No Health Consequences or Impact	A0908	Incorrect, Inadequate or Imprecise Result or Readings			
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A0908	Incorrect, Inadequate or Imprecise Result or Readings			
							F26	No Health Consequences or Impact	A0908	Incorrect, Inadequate or Imprecise Result or Readings			
					CT2727	Active implant devices/ancillaries	E2403	No Clinical Signs, Symptoms or Conditions	A0908	Incorrect, Inadequate or Imprecise Result or Readings			
							F26	No Health Consequences or Impact	A0908	Incorrect, Inadequate or Imprecise Result or Readings			
					CT2750	Implantable cardiac electrical stimulators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A0908	Incorrect, Inadequate or Imprecise Result or Readings			
							F26	No Health Consequences or Impact	A0908	Incorrect, Inadequate or Imprecise Result or Readings			
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A0908	Incorrect, Inadequate or Imprecise Result or Readings			
							F26	No Health Consequences or Impact	A0908	Incorrect, Inadequate or Imprecise Result or Readings			
					CT2985	Active	E2403	No Clinical Signs, Symptoms or Conditions	A0908	Incorrect, Inadequate or Imprecise Result or Readings			
							F26	No Health Consequences or Impact	A0908	Incorrect, Inadequate or Imprecise Result or Readings			
					CT3001	Implantable metal	E2403	No Clinical Signs, Symptoms or Conditions	A0908	Incorrect, Inadequate or Imprecise Result or Readings			
							F26	No Health Consequences or Impact	A0908	Incorrect, Inadequate or Imprecise Result or Readings			
					CT3002	Implantable synthetic polymer	E2403	No Clinical Signs, Symptoms or Conditions	A0908	Incorrect, Inadequate or Imprecise Result or Readings			
							F26	No Health Consequences or Impact	A0908	Incorrect, Inadequate or Imprecise Result or Readings			
					CT3006	Implantable inorganic material	E2403	No Clinical Signs, Symptoms or Conditions	A0908	Incorrect, Inadequate or Imprecise Result or Readings			
							F26	No Health Consequences or Impact	A0908	Incorrect, Inadequate or Imprecise Result or Readings			
					CT314	Electrically powered	E2403	No Clinical Signs, Symptoms or Conditions	A0908	Incorrect, Inadequate or Imprecise Result or Readings			
							F26	No Health Consequences or Impact	A0908	Incorrect, Inadequate or Imprecise Result or Readings			
					CT321	Long-term surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A0908	Incorrect, Inadequate or Imprecise Result or Readings			
							F26	No Health Consequences or Impact	A0908	Incorrect, Inadequate or Imprecise Result or Readings			
					CT449	Generators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A0908	Incorrect, Inadequate or Imprecise Result or Readings			
							F26	No Health Consequences or Impact	A0908	Incorrect, Inadequate or Imprecise Result or Readings			
					CT981	Single use	E2403	No Clinical Signs, Symptoms or Conditions	A0908	Incorrect, Inadequate or Imprecise Result or Readings			
							F26	No Health Consequences or Impact	A0908	Incorrect, Inadequate or Imprecise Result or Readings			
					CT983	Surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A0908	Incorrect, Inadequate or Imprecise Result or Readings			
							F26	No Health Consequences or Impact	A0908	Incorrect, Inadequate or Imprecise Result or Readings			
13329	2023	All other reportable incidents	47265	Dual-chamber implantable pacemaker, rate-responsive	CT141	Electrodes/Leads and associated devices	E0112	Dizziness	A0407	Material Discolored	02/06/2023	02/06/2023
									A0723	Unexpected Mode Switch	02/06/2023	02/06/2023
							E0601	Arrhythmia	A0407	Material Discolored	02/06/2023	02/06/2023
									A0723	Unexpected Mode Switch	02/06/2023	02/06/2023
							E2311	Discomfort	A0407	Material Discolored	02/06/2023	02/06/2023
									A0723	Unexpected Mode Switch	02/06/2023	02/06/2023
							F1901	Additional Surgery	A0407	Material Discolored	02/06/2023	02/06/2023
									A0723	Unexpected Mode Switch	02/06/2023	02/06/2023
					CT145	Cardiovascular devices	E0112	Dizziness	A0407	Material Discolored	02/06/2023	02/06/2023
									A0723	Unexpected Mode Switch	02/06/2023	02/06/2023
							E0601	Arrhythmia	A0407	Material Discolored	02/06/2023	02/06/2023
									A0723	Unexpected Mode Switch	02/06/2023	02/06/2023
							E2311	Discomfort	A0407	Material Discolored	02/06/2023	02/06/2023
									A0723	Unexpected Mode Switch	02/06/2023	02/06/2023
							F1901	Additional Surgery	A0407	Material Discolored	02/06/2023	02/06/2023
									A0723	Unexpected Mode Switch	02/06/2023	02/06/2023
					CT1684	Cardiac pacemakers and associated devices	E0112	Dizziness	A0407	Material Discolored	02/06/2023	02/06/2023
									A0723	Unexpected Mode Switch	02/06/2023	02/06/2023
							E0601	Arrhythmia	A0407	Material Discolored	02/06/2023	02/06/2023
									A0723	Unexpected Mode Switch	02/06/2023	02/06/2023
							E2311	Discomfort	A0407	Material Discolored	02/06/2023	02/06/2023
									A0723	Unexpected Mode Switch	02/06/2023	02/06/2023
							F1901	Additional Surgery	A0407	Material Discolored	02/06/2023	02/06/2023
									A0723	Unexpected Mode Switch	02/06/2023	02/06/2023
					CT1685	Cardiac pacemakers	E0112	Dizziness	A0407	Material Discolored	02/06/2023	02/06/2023
									A0723	Unexpected Mode Switch	02/06/2023	02/06/2023
							E0601	Arrhythmia	A0407	Material Discolored	02/06/2023	02/06/2023
									A0723	Unexpected Mode Switch	02/06/2023	02/06/2023
							E2311	Discomfort	A0407	Material Discolored	02/06/2023	02/06/2023
									A0723	Unexpected Mode Switch	02/06/2023	02/06/2023
							F1901	Additional Surgery	A0407	Material Discolored	02/06/2023	02/06/2023
									A0723	Unexpected Mode Switch	02/06/2023	02/06/2023
					CT233	Surgical	E0112	Dizziness	A0407	Material Discolored	02/06/2023	02/06/2023

								A0723	Unexpected Mode Switch	02/06/2023	02/06/2023
						E0601	Arrhythmia	A0407	Material Discolored	02/06/2023	02/06/2023
								A0723	Unexpected Mode Switch	02/06/2023	02/06/2023
						E2311	Discomfort	A0407	Material Discolored	02/06/2023	02/06/2023
								A0723	Unexpected Mode Switch	02/06/2023	02/06/2023
						F1901	Additional Surgery	A0407	Material Discolored	02/06/2023	02/06/2023
								A0723	Unexpected Mode Switch	02/06/2023	02/06/2023
					CT2406	Implantable	E0112	Dizziness	A0407	Material Discolored	02/06/2023	02/06/2023
									A0723	Unexpected Mode Switch	02/06/2023	02/06/2023
							E0601	Arrhythmia	A0407	Material Discolored	02/06/2023	02/06/2023
									A0723	Unexpected Mode Switch	02/06/2023	02/06/2023
							E2311	Discomfort	A0407	Material Discolored	02/06/2023	02/06/2023
									A0723	Unexpected Mode Switch	02/06/2023	02/06/2023
							F1901	Additional Surgery	A0407	Material Discolored	02/06/2023	02/06/2023
									A0723	Unexpected Mode Switch	02/06/2023	02/06/2023
					CT2407	Active implantable	E0112	Dizziness	A0407	Material Discolored	02/06/2023	02/06/2023
									A0723	Unexpected Mode Switch	02/06/2023	02/06/2023
							E0601	Arrhythmia	A0407	Material Discolored	02/06/2023	02/06/2023
									A0723	Unexpected Mode Switch	02/06/2023	02/06/2023
							E2311	Discomfort	A0407	Material Discolored	02/06/2023	02/06/2023
									A0723	Unexpected Mode Switch	02/06/2023	02/06/2023
							F1901	Additional Surgery	A0407	Material Discolored	02/06/2023	02/06/2023
									A0723	Unexpected Mode Switch	02/06/2023	02/06/2023
					CT257	Implantable cardiac pacemakers	E0112	Dizziness	A0407	Material Discolored	02/06/2023	02/06/2023
									A0723	Unexpected Mode Switch	02/06/2023	02/06/2023
							E0601	Arrhythmia	A0407	Material Discolored	02/06/2023	02/06/2023
									A0723	Unexpected Mode Switch	02/06/2023	02/06/2023
							E2311	Discomfort	A0407	Material Discolored	02/06/2023	02/06/2023
									A0723	Unexpected Mode Switch	02/06/2023	02/06/2023
							F1901	Additional Surgery	A0407	Material Discolored	02/06/2023	02/06/2023
									A0723	Unexpected Mode Switch	02/06/2023	02/06/2023
					CT2727	Active implant devices/ancillaries	E0112	Dizziness	A0407	Material Discolored	02/06/2023	02/06/2023
									A0723	Unexpected Mode Switch	02/06/2023	02/06/2023
							E0601	Arrhythmia	A0407	Material Discolored	02/06/2023	02/06/2023
									A0723	Unexpected Mode Switch	02/06/2023	02/06/2023
							E2311	Discomfort	A0407	Material Discolored	02/06/2023	02/06/2023
									A0723	Unexpected Mode Switch	02/06/2023	02/06/2023
							F1901	Additional Surgery	A0407	Material Discolored	02/06/2023	02/06/2023
									A0723	Unexpected Mode Switch	02/06/2023	02/06/2023
					CT2750	Implantable cardiac electrical stimulators and associated devices	E0112	Dizziness	A0407	Material Discolored	02/06/2023	02/06/2023
									A0723	Unexpected Mode Switch	02/06/2023	02/06/2023
							E0601	Arrhythmia	A0407	Material Discolored	02/06/2023	02/06/2023
									A0723	Unexpected Mode Switch	02/06/2023	02/06/2023
							E2311	Discomfort	A0407	Material Discolored	02/06/2023	02/06/2023
									A0723	Unexpected Mode Switch	02/06/2023	02/06/2023
							F1901	Additional Surgery	A0407	Material Discolored	02/06/2023	02/06/2023
									A0723	Unexpected Mode Switch	02/06/2023	02/06/2023
					CT2866	Implantable cardiac electrical stimulators	E0112	Dizziness	A0407	Material Discolored	02/06/2023	02/06/2023
									A0723	Unexpected Mode Switch	02/06/2023	02/06/2023
							E0601	Arrhythmia	A0407	Material Discolored	02/06/2023	02/06/2023
									A0723	Unexpected Mode Switch	02/06/2023	02/06/2023
							E2311	Discomfort	A0407	Material Discolored	02/06/2023	02/06/2023
									A0723	Unexpected Mode Switch	02/06/2023	02/06/2023
							F1901	Additional Surgery	A0407	Material Discolored	02/06/2023	02/06/2023
									A0723	Unexpected Mode Switch	02/06/2023	02/06/2023
					CT2985	Active	E0112	Dizziness	A0407	Material Discolored	02/06/2023	02/06/2023
									A0723	Unexpected Mode Switch	02/06/2023	02/06/2023
							E0601	Arrhythmia	A0407	Material Discolored	02/06/2023	02/06/2023
									A0723	Unexpected Mode Switch	02/06/2023	02/06/2023
							E2311	Discomfort	A0407	Material Discolored	02/06/2023	02/06/2023
									A0723	Unexpected Mode Switch	02/06/2023	02/06/2023
							F1901	Additional Surgery	A0407	Material Discolored	02/06/2023	02/06/2023
									A0723	Unexpected Mode Switch	02/06/2023	02/06/2023
					CT3001	Implantable metal	E0112	Dizziness	A0407	Material Discolored	02/06/2023	02/06/2023
									A0723	Unexpected Mode Switch	02/06/2023	02/06/2023
							E0601	Arrhythmia	A0407	Material Discolored	02/06/2023	02/06/2023
									A0723	Unexpected Mode Switch	02/06/2023	02/06/2023
							E2311	Discomfort	A0407	Material Discolored	02/06/2023	02/06/2023
									A0723	Unexpected Mode Switch	02/06/2023	02/06/2023
							F1901	Additional Surgery	A0407	Material Discolored	02/06/2023	02/06/2023
									A0723	Unexpected Mode Switch	02/06/2023	02/06/2023
					CT3002	Implantable synthetic polymer	E0112	Dizziness	A0407	Material Discolored	02/06/2023	02/06/2023
									A0723	Unexpected Mode Switch	02/06/2023	02/06/2023
							E0601	Arrhythmia	A0407	Material Discolored	02/06/2023	02/06/2023
									A0723	Unexpected Mode Switch	02/06/2023	02/06/2023
							E2311	Discomfort	A0407	Material Discolored	02/06/2023	02/06/2023
									A0723	Unexpected Mode Switch	02/06/2023	02/06/2023
							F1901	Additional Surgery	A0407	Material Discolored	02/06/2023	02/06/2023
									A0723	Unexpected Mode Switch	02/06/2023	02/06/2023
					CT3006	Implantable inorganic material	E0112	Dizziness	A0407	Material Discolored	02/06/2023	02/06/2023
									A0723	Unexpected Mode Switch	02/06/2023	02/06/2023
							E0601	Arrhythmia	A0407	Material Discolored	02/06/2023	02/06/2023
									A0723	Unexpected Mode Switch	02/06/2023	02/06/2023
							E2311	Discomfort	A0407	Material Discolored	02/06/2023	02/06/2023
									A0723	Unexpected Mode Switch	02/06/2023	02/06/2023
							F1901	Additional Surgery	A0407	Material Discolored	02/06/2023	02/06/2023
									A0723	Unexpected Mode Switch	02/06/2023	02/06/2023
					CT314	Electrically powered	E0112	Dizziness	A0407	Material Discolored	02/06/2023	02/06/2023

									A0723	Unexpected Mode Switch	02/06/2023	02/06/2023
							E0601	Arrhythmia	A0407	Material Discolored	02/06/2023	02/06/2023
									A0723	Unexpected Mode Switch	02/06/2023	02/06/2023
							E2311	Discomfort	A0407	Material Discolored	02/06/2023	02/06/2023
									A0723	Unexpected Mode Switch	02/06/2023	02/06/2023
							F1901	Additional Surgery	A0407	Material Discolored	02/06/2023	02/06/2023
									A0723	Unexpected Mode Switch	02/06/2023	02/06/2023
					CT321	Long-term surgical invasive	E0112	Dizziness	A0407	Material Discolored	02/06/2023	02/06/2023
									A0723	Unexpected Mode Switch	02/06/2023	02/06/2023
							E0601	Arrhythmia	A0407	Material Discolored	02/06/2023	02/06/2023
									A0723	Unexpected Mode Switch	02/06/2023	02/06/2023
							E2311	Discomfort	A0407	Material Discolored	02/06/2023	02/06/2023
									A0723	Unexpected Mode Switch	02/06/2023	02/06/2023
							F1901	Additional Surgery	A0407	Material Discolored	02/06/2023	02/06/2023
									A0723	Unexpected Mode Switch	02/06/2023	02/06/2023
					CT449	Generators and associated devices	E0112	Dizziness	A0407	Material Discolored	02/06/2023	02/06/2023
									A0723	Unexpected Mode Switch	02/06/2023	02/06/2023
							E0601	Arrhythmia	A0407	Material Discolored	02/06/2023	02/06/2023
									A0723	Unexpected Mode Switch	02/06/2023	02/06/2023
							E2311	Discomfort	A0407	Material Discolored	02/06/2023	02/06/2023
									A0723	Unexpected Mode Switch	02/06/2023	02/06/2023
							F1901	Additional Surgery	A0407	Material Discolored	02/06/2023	02/06/2023
									A0723	Unexpected Mode Switch	02/06/2023	02/06/2023
					CT981	Single use	E0112	Dizziness	A0407	Material Discolored	02/06/2023	02/06/2023
									A0723	Unexpected Mode Switch	02/06/2023	02/06/2023
							E0601	Arrhythmia	A0407	Material Discolored	02/06/2023	02/06/2023
									A0723	Unexpected Mode Switch	02/06/2023	02/06/2023
							E2311	Discomfort	A0407	Material Discolored	02/06/2023	02/06/2023
									A0723	Unexpected Mode Switch	02/06/2023	02/06/2023
							F1901	Additional Surgery	A0407	Material Discolored	02/06/2023	02/06/2023
									A0723	Unexpected Mode Switch	02/06/2023	02/06/2023
					CT983	Surgical invasive	E0112	Dizziness	A0407	Material Discolored	02/06/2023	02/06/2023
									A0723	Unexpected Mode Switch	02/06/2023	02/06/2023
							E0601	Arrhythmia	A0407	Material Discolored	02/06/2023	02/06/2023
									A0723	Unexpected Mode Switch	02/06/2023	02/06/2023
							E2311	Discomfort	A0407	Material Discolored	02/06/2023	02/06/2023
									A0723	Unexpected Mode Switch	02/06/2023	02/06/2023
							F1901	Additional Surgery	A0407	Material Discolored	02/06/2023	02/06/2023
									A0723	Unexpected Mode Switch	02/06/2023	02/06/2023
13330	2023		47270	Cardiac resynchronization therapy implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E0601	Arrhythmia	A070910	Undersensing	05/06/2023	05/06/2023
							F02	Death	A070910	Undersensing	05/06/2023	05/06/2023
13331	2023	All other reportable incidents	35223	Endocardial/interventricular septal pacing lead	CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A0401	Break	25/05/2023	25/05/2023
									A070101	Failure to Capture	25/05/2023	25/05/2023
									A070908	Failure to Sense	25/05/2023	25/05/2023
									A072202	Low Impedance	25/05/2023	25/05/2023
							F1908	Prolonged Surgery	A0401	Break	25/05/2023	25/05/2023
									A070101	Failure to Capture	25/05/2023	25/05/2023
									A070908	Failure to Sense	25/05/2023	25/05/2023
									A072202	Low Impedance	25/05/2023	25/05/2023
13332	2023		47270	Cardiac resynchronization therapy implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E0112	Dizziness	A070909	Oversensing	01/06/2023	01/06/2023
									A071204	No Pacing	01/06/2023	01/06/2023
							F1901	Additional Surgery	A070909	Oversensing	01/06/2023	01/06/2023
									A071204	No Pacing	01/06/2023	01/06/2023
13333	2023	All other reportable incidents	17882	Physiologic-monitoring defibrillation system	CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A0719	Unexpected Shutdown		
							F26	No Health Consequences or Impact	A0719	Unexpected Shutdown		
13334	2023	All other reportable incidents	37265	Dual-chamber implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A1601	Device Alarm System		
							F26	No Health Consequences or Impact	A1601	Device Alarm System		
13335	2023		47265	Dual-chamber implantable pacemaker, rate-responsive	CT141	Electrodes/Leads and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	04/05/2023	04/05/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	04/05/2023	04/05/2023
					CT145	Cardiovascular devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	04/05/2023	04/05/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	04/05/2023	04/05/2023
					CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	04/05/2023	04/05/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	04/05/2023	04/05/2023
					CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	04/05/2023	04/05/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	04/05/2023	04/05/2023
					CT233	Surgical	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	04/05/2023	04/05/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	04/05/2023	04/05/2023
					CT2406	Implantable	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	04/05/2023	04/05/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	04/05/2023	04/05/2023
					CT2407	Active implantable	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	04/05/2023	04/05/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	04/05/2023	04/05/2023
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	04/05/2023	04/05/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	04/05/2023	04/05/2023
					CT2727	Active implant devices/ancillaries	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	04/05/2023	04/05/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	04/05/2023	04/05/2023
					CT2750	Implantable cardiac electrical stimulators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	04/05/2023	04/05/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	04/05/2023	04/05/2023
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	04/05/2023	04/05/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	04/05/2023	04/05/2023
					CT2985	Active	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	04/05/2023	04/05/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	04/05/2023	04/05/2023
					CT3001	Implantable metal	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	04/05/2023	04/05/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	04/05/2023	04/05/2023
					CT3002	Implantable synthetic polymer	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	04/05/2023	04/05/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	04/05/2023	04/05/2023
					CT3006	Implantable inorganic material	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	04/05/2023	04/05/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	04/05/2023	04/05/2023
					CT314	Electrically powered	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	04/05/2023	04/05/2023

							F1901	Additional Surgery	A24	No Apparent Device Problem	04/05/2023	04/05/2023
					CT321	Long-term surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	04/05/2023	04/05/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	04/05/2023	04/05/2023
					CT449	Generators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	04/05/2023	04/05/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	04/05/2023	04/05/2023
					CT981	Single use	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	04/05/2023	04/05/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	04/05/2023	04/05/2023
					CT983	Surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	04/05/2023	04/05/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	04/05/2023	04/05/2023
13336	2023		47265	Dual-chamber implantable pacemaker, rate-responsive	CT141	Electrodes/Leads and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	27/04/2023	27/04/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	27/04/2023	27/04/2023
					CT145	Cardiovascular devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	27/04/2023	27/04/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	27/04/2023	27/04/2023
					CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	27/04/2023	27/04/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	27/04/2023	27/04/2023
					CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	27/04/2023	27/04/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	27/04/2023	27/04/2023
					CT233	Surgical	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	27/04/2023	27/04/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	27/04/2023	27/04/2023
					CT2406	Implantable	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	27/04/2023	27/04/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	27/04/2023	27/04/2023
					CT2407	Active implantable	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	27/04/2023	27/04/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	27/04/2023	27/04/2023
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	27/04/2023	27/04/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	27/04/2023	27/04/2023
					CT2727	Active implant devices/ancillaries	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	27/04/2023	27/04/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	27/04/2023	27/04/2023
					CT2750	Implantable cardiac electrical stimulators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	27/04/2023	27/04/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	27/04/2023	27/04/2023
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	27/04/2023	27/04/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	27/04/2023	27/04/2023
					CT2985	Active	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	27/04/2023	27/04/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	27/04/2023	27/04/2023
					CT3001	Implantable metal	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	27/04/2023	27/04/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	27/04/2023	27/04/2023
					CT3002	Implantable synthetic polymer	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	27/04/2023	27/04/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	27/04/2023	27/04/2023
					CT3006	Implantable inorganic material	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	27/04/2023	27/04/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	27/04/2023	27/04/2023
					CT314	Electrically powered	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	27/04/2023	27/04/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	27/04/2023	27/04/2023
					CT321	Long-term surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	27/04/2023	27/04/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	27/04/2023	27/04/2023
					CT449	Generators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	27/04/2023	27/04/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	27/04/2023	27/04/2023
					CT981	Single use	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	27/04/2023	27/04/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	27/04/2023	27/04/2023
					CT983	Surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	27/04/2023	27/04/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	27/04/2023	27/04/2023
13337	2023	All other reportable incidents	47265	Dual-chamber implantable pacemaker, rate-responsive	CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A0722	Impedance Problem		
							F26	No Health Consequences or Impact	A0722	Impedance Problem		
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A0722	Impedance Problem		
							F26	No Health Consequences or Impact	A0722	Impedance Problem		
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A0722	Impedance Problem		
							F26	No Health Consequences or Impact	A0722	Impedance Problem		
13338	2023	All other reportable incidents	47265	Dual-chamber implantable pacemaker, rate-responsive	CT141	Electrodes/Leads and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A0722	Impedance Problem		
							F26	No Health Consequences or Impact	A0722	Impedance Problem		
					CT145	Cardiovascular devices	E2403	No Clinical Signs, Symptoms or Conditions	A0722	Impedance Problem		
							F26	No Health Consequences or Impact	A0722	Impedance Problem		
					CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A0722	Impedance Problem		
							F26	No Health Consequences or Impact	A0722	Impedance Problem		
					CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A0722	Impedance Problem		
							F26	No Health Consequences or Impact	A0722	Impedance Problem		
					CT233	Surgical	E2403	No Clinical Signs, Symptoms or Conditions	A0722	Impedance Problem		
							F26	No Health Consequences or Impact	A0722	Impedance Problem		
					CT2406	Implantable	E2403	No Clinical Signs, Symptoms or Conditions	A0722	Impedance Problem		
							F26	No Health Consequences or Impact	A0722	Impedance Problem		
					CT2407	Active implantable	E2403	No Clinical Signs, Symptoms or Conditions	A0722	Impedance Problem		
							F26	No Health Consequences or Impact	A0722	Impedance Problem		
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A0722	Impedance Problem		
							F26	No Health Consequences or Impact	A0722	Impedance Problem		
					CT2727	Active implant devices/ancillaries	E2403	No Clinical Signs, Symptoms or Conditions	A0722	Impedance Problem		
							F26	No Health Consequences or Impact	A0722	Impedance Problem		
					CT2750	Implantable cardiac electrical stimulators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A0722	Impedance Problem		
							F26	No Health Consequences or Impact	A0722	Impedance Problem		
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A0722	Impedance Problem		
							F26	No Health Consequences or Impact	A0722	Impedance Problem		
					CT2985	Active	E2403	No Clinical Signs, Symptoms or Conditions	A0722	Impedance Problem		
							F26	No Health Consequences or Impact	A0722	Impedance Problem		
					CT3001	Implantable metal	E2403	No Clinical Signs, Symptoms or Conditions	A0722	Impedance Problem		
							F26	No Health Consequences or Impact	A0722	Impedance Problem		
					CT3002	Implantable synthetic polymer	E2403	No Clinical Signs, Symptoms or Conditions	A0722	Impedance Problem		
							F26	No Health Consequences or Impact	A0722	Impedance Problem		
					CT3006	Implantable inorganic material	E2403	No Clinical Signs, Symptoms or Conditions	A0722	Impedance Problem		
							F26	No Health Consequences or Impact	A0722	Impedance Problem		
					CT314	Electrically powered	E2403	No Clinical Signs, Symptoms or Conditions	A0722	Impedance Problem		
							F26	No Health Consequences or Impact	A0722	Impedance Problem		
					CT321	Long-term surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A0722	Impedance Problem		

						F26	No Health Consequences or Impact	A0722	Impedance Problem			
				CT449	Generators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A0722	Impedance Problem			
						F26	No Health Consequences or Impact	A0722	Impedance Problem			
				CT981	Single use	E2403	No Clinical Signs, Symptoms or Conditions	A0722	Impedance Problem			
						F26	No Health Consequences or Impact	A0722	Impedance Problem			
				CT983	Surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A0722	Impedance Problem			
						F26	No Health Consequences or Impact	A0722	Impedance Problem			
13339	2023	All other reportable incidents	47265	Dual-chamber implantable pacemaker, rate-responsive	CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A0722	Impedance Problem		
							F26	No Health Consequences or Impact	A0722	Impedance Problem		
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A0722	Impedance Problem		
							F26	No Health Consequences or Impact	A0722	Impedance Problem		
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A0722	Impedance Problem		
							F26	No Health Consequences or Impact	A0722	Impedance Problem		
13340	2023		47265	Dual-chamber implantable pacemaker, rate-responsive	CT141	Electrodes/Leads and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	08/06/2023	08/06/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	08/06/2023	08/06/2023
					CT145	Cardiovascular devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	08/06/2023	08/06/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	08/06/2023	08/06/2023
					CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	08/06/2023	08/06/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	08/06/2023	08/06/2023
					CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	08/06/2023	08/06/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	08/06/2023	08/06/2023
					CT233	Surgical	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	08/06/2023	08/06/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	08/06/2023	08/06/2023
					CT2406	Implantable	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	08/06/2023	08/06/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	08/06/2023	08/06/2023
					CT2407	Active implantable	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	08/06/2023	08/06/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	08/06/2023	08/06/2023
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	08/06/2023	08/06/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	08/06/2023	08/06/2023
					CT2727	Active implant devices/ancillaries	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	08/06/2023	08/06/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	08/06/2023	08/06/2023
					CT2750	Implantable cardiac electrical stimulators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	08/06/2023	08/06/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	08/06/2023	08/06/2023
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	08/06/2023	08/06/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	08/06/2023	08/06/2023
					CT2985	Active	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	08/06/2023	08/06/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	08/06/2023	08/06/2023
					CT3001	Implantable metal	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	08/06/2023	08/06/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	08/06/2023	08/06/2023
					CT3002	Implantable synthetic polymer	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	08/06/2023	08/06/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	08/06/2023	08/06/2023
					CT3006	Implantable inorganic material	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	08/06/2023	08/06/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	08/06/2023	08/06/2023
					CT314	Electrically powered	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	08/06/2023	08/06/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	08/06/2023	08/06/2023
					CT321	Long-term surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	08/06/2023	08/06/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	08/06/2023	08/06/2023
					CT449	Generators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	08/06/2023	08/06/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	08/06/2023	08/06/2023
					CT981	Single use	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	08/06/2023	08/06/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	08/06/2023	08/06/2023
					CT983	Surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	08/06/2023	08/06/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	08/06/2023	08/06/2023
13341	2023		35224	External pacemaker, epicardial pacing	CT1684	Cardiac pacemakers and associated devices	F26	No Health Consequences or Impact	A02	Manufacturing, Packaging or Shipping Problem		
					CT1685	Cardiac pacemakers	F26	No Health Consequences or Impact	A02	Manufacturing, Packaging or Shipping Problem		
13342	2023	All other reportable incidents	47205	Cardiac pulse generator programmer	CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A0203	Defective Device		
									A150105	Key or Button Unresponsive/not Working		
							F26	No Health Consequences or Impact	A0203	Defective Device		
									A150105	Key or Button Unresponsive/not Working		
13343	2023		17882	Physiologic-monitoring defibrillation system	CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A0201	Product Quality Problem		
									A0712	Pacing Problem		
							F26	No Health Consequences or Impact	A0201	Product Quality Problem		
									A0712	Pacing Problem		
13344	2023	All other reportable incidents	35223	Endocardial/interventricular septal pacing lead	CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A090801	Signal Artifact/Noise		
							F1901	Additional Surgery	A090801	Signal Artifact/Noise		
13345	2023	All other reportable incidents	47265	Dual-chamber implantable pacemaker, rate-responsive	CT141	Electrodes/Leads and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	01/06/2023	01/06/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	01/06/2023	01/06/2023
					CT145	Cardiovascular devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	01/06/2023	01/06/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	01/06/2023	01/06/2023
					CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	01/06/2023	01/06/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	01/06/2023	01/06/2023
					CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	01/06/2023	01/06/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	01/06/2023	01/06/2023
					CT233	Surgical	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	01/06/2023	01/06/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	01/06/2023	01/06/2023
					CT2406	Implantable	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	01/06/2023	01/06/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	01/06/2023	01/06/2023
					CT2407	Active implantable	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	01/06/2023	01/06/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	01/06/2023	01/06/2023
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	01/06/2023	01/06/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	01/06/2023	01/06/2023
					CT2727	Active implant devices/ancillaries	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	01/06/2023	01/06/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	01/06/2023	01/06/2023
					CT2750	Implantable cardiac electrical stimulators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	01/06/2023	01/06/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	01/06/2023	01/06/2023
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	01/06/2023	01/06/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	01/06/2023	01/06/2023
					CT2985	Active	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	01/06/2023	01/06/2023

							F1901	Additional Surgery	A24	No Apparent Device Problem	01/06/2023	01/06/2023
					CT3001	Implantable metal	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	01/06/2023	01/06/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	01/06/2023	01/06/2023
					CT3002	Implantable synthetic polymer	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	01/06/2023	01/06/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	01/06/2023	01/06/2023
					CT3006	Implantable inorganic material	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	01/06/2023	01/06/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	01/06/2023	01/06/2023
					CT314	Electrically powered	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	01/06/2023	01/06/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	01/06/2023	01/06/2023
					CT321	Long-term surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	01/06/2023	01/06/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	01/06/2023	01/06/2023
					CT449	Generators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	01/06/2023	01/06/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	01/06/2023	01/06/2023
					CT981	Single use	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	01/06/2023	01/06/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	01/06/2023	01/06/2023
					CT983	Surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	01/06/2023	01/06/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	01/06/2023	01/06/2023
13346	2023	All other reportable incidents	35223	Endocardial/interventricular septal pacing lead	CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A0401	Break		
									A070909	Oversensing		
							F26	No Health Consequences or Impact	A0401	Break		
									A070909	Oversensing		
13347	2023	All other reportable incidents	47265	Dual-chamber implantable pacemaker, rate-responsive	CT141	Electrodes/Leads and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	05/05/2023	05/05/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	05/05/2023	05/05/2023
					CT145	Cardiovascular devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	05/05/2023	05/05/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	05/05/2023	05/05/2023
					CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	05/05/2023	05/05/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	05/05/2023	05/05/2023
					CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	05/05/2023	05/05/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	05/05/2023	05/05/2023
					CT233	Surgical	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	05/05/2023	05/05/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	05/05/2023	05/05/2023
					CT2406	Implantable	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	05/05/2023	05/05/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	05/05/2023	05/05/2023
					CT2407	Active implantable	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	05/05/2023	05/05/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	05/05/2023	05/05/2023
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	05/05/2023	05/05/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	05/05/2023	05/05/2023
					CT2727	Active implant devices/ancillaries	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	05/05/2023	05/05/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	05/05/2023	05/05/2023
					CT2750	Implantable cardiac electrical stimulators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	05/05/2023	05/05/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	05/05/2023	05/05/2023
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	05/05/2023	05/05/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	05/05/2023	05/05/2023
					CT2985	Active	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	05/05/2023	05/05/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	05/05/2023	05/05/2023
					CT3001	Implantable metal	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	05/05/2023	05/05/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	05/05/2023	05/05/2023
					CT3002	Implantable synthetic polymer	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	05/05/2023	05/05/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	05/05/2023	05/05/2023
					CT3006	Implantable inorganic material	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	05/05/2023	05/05/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	05/05/2023	05/05/2023
					CT314	Electrically powered	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	05/05/2023	05/05/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	05/05/2023	05/05/2023
					CT321	Long-term surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	05/05/2023	05/05/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	05/05/2023	05/05/2023
					CT449	Generators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	05/05/2023	05/05/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	05/05/2023	05/05/2023
					CT981	Single use	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	05/05/2023	05/05/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	05/05/2023	05/05/2023
					CT983	Surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	05/05/2023	05/05/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	05/05/2023	05/05/2023
13348	2023		47265	Dual-chamber implantable pacemaker, rate-responsive	CT141	Electrodes/Leads and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	08/03/2023	08/03/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	08/03/2023	08/03/2023
					CT145	Cardiovascular devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	08/03/2023	08/03/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	08/03/2023	08/03/2023
					CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	08/03/2023	08/03/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	08/03/2023	08/03/2023
					CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	08/03/2023	08/03/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	08/03/2023	08/03/2023
					CT233	Surgical	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	08/03/2023	08/03/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	08/03/2023	08/03/2023
					CT2406	Implantable	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	08/03/2023	08/03/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	08/03/2023	08/03/2023
					CT2407	Active implantable	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	08/03/2023	08/03/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	08/03/2023	08/03/2023
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	08/03/2023	08/03/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	08/03/2023	08/03/2023
					CT2727	Active implant devices/ancillaries	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	08/03/2023	08/03/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	08/03/2023	08/03/2023
					CT2750	Implantable cardiac electrical stimulators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	08/03/2023	08/03/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	08/03/2023	08/03/2023
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	08/03/2023	08/03/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	08/03/2023	08/03/2023
					CT2985	Active	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	08/03/2023	08/03/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	08/03/2023	08/03/2023
					CT3001	Implantable metal	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	08/03/2023	08/03/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	08/03/2023	08/03/2023
					CT3002	Implantable synthetic polymer	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	08/03/2023	08/03/2023

						F1901	Additional Surgery	A24	No Apparent Device Problem	08/03/2023	08/03/2023
					CT3006	Implantable inorganic material	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	08/03/2023	08/03/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	08/03/2023	08/03/2023
					CT314	Electrically powered	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	08/03/2023	08/03/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	08/03/2023	08/03/2023
					CT321	Long-term surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	08/03/2023	08/03/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	08/03/2023	08/03/2023
					CT449	Generators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	08/03/2023	08/03/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	08/03/2023	08/03/2023
					CT981	Single use	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	08/03/2023	08/03/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	08/03/2023	08/03/2023
					CT983	Surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	08/03/2023	08/03/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	08/03/2023	08/03/2023
13349	2023	All other reportable incidents	47267	Single-chamber implantable pacemaker, rate-responsive	CT141	Electrodes/Leads and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
					CT145	Cardiovascular devices	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
					CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
					CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
					CT233	Surgical	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
					CT2406	Implantable	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
					CT2407	Active implantable	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
					CT2727	Active implant devices/ancillaries	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
					CT2750	Implantable cardiac electrical stimulators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
					CT2985	Active	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
					CT3001	Implantable metal	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
					CT3002	Implantable synthetic polymer	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
					CT3006	Implantable inorganic material	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
					CT314	Electrically powered	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
					CT321	Long-term surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
					CT449	Generators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
					CT981	Single use	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A0705	Battery Problem		

									A070504	Premature Discharge of Battery		
				CT983	Surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem			
								A070504	Premature Discharge of Battery			
						F26	No Health Consequences or Impact	A0705	Battery Problem			
								A070504	Premature Discharge of Battery			
13350	2023	All other reportable incidents	47270	Cardiac resynchronization therapy implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A0712	Pacing Problem		
									A1301	Failure to Read Input Signal		
							F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A0712	Pacing Problem		
									A1301	Failure to Read Input Signal		
13351	2023	All other reportable incidents	47267	Single-chamber implantable pacemaker, rate-responsive	CT141	Electrodes/Leads and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT145	Cardiovascular devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT233	Surgical	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT2406	Implantable	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT2407	Active implantable	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT2727	Active implant devices/ancillaries	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT2750	Implantable cardiac electrical stimulators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT2985	Active	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT3001	Implantable metal	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT3002	Implantable synthetic polymer	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT3006	Implantable inorganic material	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT314	Electrically powered	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT321	Long-term surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT449	Generators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT981	Single use	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT983	Surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
13352	2023	All other reportable incidents	17882	Physiologic-monitoring defibrillation system	CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A0719	Unexpected Shutdown		
							F27	Problem Identified During Non-Clinical Procedure	A0719	Unexpected Shutdown		
13353	2023	All other reportable incidents	47263	Cardiac resynchronization therapy implantable pacemaker	CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A0908	Incorrect, Inadequate or Imprecise Result or Readings	27/02/2023	27/02/2023
							F08	Hospitalization or Prolonged Hospitalization	A0908	Incorrect, Inadequate or Imprecise Result or Readings	27/02/2023	27/02/2023
							F1905	Device Revision or Replacement	A0908	Incorrect, Inadequate or Imprecise Result or Readings	27/02/2023	27/02/2023
					CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A0908	Incorrect, Inadequate or Imprecise Result or Readings	27/02/2023	27/02/2023
							F08	Hospitalization or Prolonged Hospitalization	A0908	Incorrect, Inadequate or Imprecise Result or Readings	27/02/2023	27/02/2023
							F1905	Device Revision or Replacement	A0908	Incorrect, Inadequate or Imprecise Result or Readings	27/02/2023	27/02/2023
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A0908	Incorrect, Inadequate or Imprecise Result or Readings	27/02/2023	27/02/2023
							F08	Hospitalization or Prolonged Hospitalization	A0908	Incorrect, Inadequate or Imprecise Result or Readings	27/02/2023	27/02/2023
							F1905	Device Revision or Replacement	A0908	Incorrect, Inadequate or Imprecise Result or Readings	27/02/2023	27/02/2023
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A0908	Incorrect, Inadequate or Imprecise Result or Readings	27/02/2023	27/02/2023
							F08	Hospitalization or Prolonged Hospitalization	A0908	Incorrect, Inadequate or Imprecise Result or Readings	27/02/2023	27/02/2023
							F1905	Device Revision or Replacement	A0908	Incorrect, Inadequate or Imprecise Result or Readings	27/02/2023	27/02/2023
13354	2023	All other reportable incidents	47263	Cardiac resynchronization therapy implantable pacemaker	CT1684	Cardiac pacemakers and associated devices	E011903	Syncope/Fainting	A070101	Failure to Capture	14/03/2023	14/03/2023
									A0908	Incorrect, Inadequate or Imprecise Result or Readings	14/03/2023	14/03/2023
							F08	Hospitalization or Prolonged Hospitalization	A070101	Failure to Capture	14/03/2023	14/03/2023
									A0908	Incorrect, Inadequate or Imprecise Result or Readings	14/03/2023	14/03/2023
							F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A070101	Failure to Capture	14/03/2023	14/03/2023
									A0908	Incorrect, Inadequate or Imprecise Result or Readings	14/03/2023	14/03/2023
							F1905	Device Revision or Replacement	A070101	Failure to Capture	14/03/2023	14/03/2023
									A0908	Incorrect, Inadequate or Imprecise Result or Readings	14/03/2023	14/03/2023
					CT1685	Cardiac pacemakers	E011903	Syncope/Fainting	A070101	Failure to Capture	14/03/2023	14/03/2023
									A0908	Incorrect, Inadequate or Imprecise Result or Readings	14/03/2023	14/03/2023
							F08	Hospitalization or Prolonged Hospitalization	A070101	Failure to Capture	14/03/2023	14/03/2023

									A0908	Incorrect, Inadequate or Imprecise Result or Readings	14/03/2023	14/03/2023
							F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A070101	Failure to Capture	14/03/2023	14/03/2023
									A0908	Incorrect, Inadequate or Imprecise Result or Readings	14/03/2023	14/03/2023
							F1905	Device Revision or Replacement	A070101	Failure to Capture	14/03/2023	14/03/2023
									A0908	Incorrect, Inadequate or Imprecise Result or Readings	14/03/2023	14/03/2023
					CT257	Implantable cardiac pacemakers	E011903	Syncope/Fainting	A070101	Failure to Capture	14/03/2023	14/03/2023
									A0908	Incorrect, Inadequate or Imprecise Result or Readings	14/03/2023	14/03/2023
							F08	Hospitalization or Prolonged Hospitalization	A070101	Failure to Capture	14/03/2023	14/03/2023
									A0908	Incorrect, Inadequate or Imprecise Result or Readings	14/03/2023	14/03/2023
							F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A070101	Failure to Capture	14/03/2023	14/03/2023
									A0908	Incorrect, Inadequate or Imprecise Result or Readings	14/03/2023	14/03/2023
							F1905	Device Revision or Replacement	A070101	Failure to Capture	14/03/2023	14/03/2023
									A0908	Incorrect, Inadequate or Imprecise Result or Readings	14/03/2023	14/03/2023
					CT2866	Implantable cardiac electrical stimulators	E011903	Syncope/Fainting	A070101	Failure to Capture	14/03/2023	14/03/2023
									A0908	Incorrect, Inadequate or Imprecise Result or Readings	14/03/2023	14/03/2023
							F08	Hospitalization or Prolonged Hospitalization	A070101	Failure to Capture	14/03/2023	14/03/2023
									A0908	Incorrect, Inadequate or Imprecise Result or Readings	14/03/2023	14/03/2023
							F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A070101	Failure to Capture	14/03/2023	14/03/2023
									A0908	Incorrect, Inadequate or Imprecise Result or Readings	14/03/2023	14/03/2023
							F1905	Device Revision or Replacement	A070101	Failure to Capture	14/03/2023	14/03/2023
									A0908	Incorrect, Inadequate or Imprecise Result or Readings	14/03/2023	14/03/2023
13355	2023	All other reportable incidents	47270	Cardiac resynchronization therapy implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A072201	High Impedance		
							F26	No Health Consequences or Impact	A072201	High Impedance		
13356	2023	All other reportable incidents	36241	Pacemaker/defibrillator lead	CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A072201	High Impedance		
							F26	No Health Consequences or Impact	A072201	High Impedance		
13357	2023	All other reportable incidents	47265	Dual-chamber implantable pacemaker, rate-responsive	CT141	Electrodes/Leads and associated devices	E0306	Sepsis	A070101	Failure to Capture		
									A0712	Pacing Problem		
									A1301	Failure to Read Input Signal		
							F08	Hospitalization or Prolonged Hospitalization	A070101	Failure to Capture		
									A0712	Pacing Problem		
									A1301	Failure to Read Input Signal		
					CT145	Cardiovascular devices	E0306	Sepsis	A070101	Failure to Capture		
									A0712	Pacing Problem		
									A1301	Failure to Read Input Signal		
							F08	Hospitalization or Prolonged Hospitalization	A070101	Failure to Capture		
									A0712	Pacing Problem		
									A1301	Failure to Read Input Signal		
					CT1684	Cardiac pacemakers and associated devices	E0306	Sepsis	A070101	Failure to Capture		
									A0712	Pacing Problem		
									A1301	Failure to Read Input Signal		
							F08	Hospitalization or Prolonged Hospitalization	A070101	Failure to Capture		
									A0712	Pacing Problem		
									A1301	Failure to Read Input Signal		
					CT1685	Cardiac pacemakers	E0306	Sepsis	A070101	Failure to Capture		
									A0712	Pacing Problem		
									A1301	Failure to Read Input Signal		
							F08	Hospitalization or Prolonged Hospitalization	A070101	Failure to Capture		
									A0712	Pacing Problem		
									A1301	Failure to Read Input Signal		
					CT233	Surgical	E0306	Sepsis	A070101	Failure to Capture		
									A0712	Pacing Problem		
									A1301	Failure to Read Input Signal		
							F08	Hospitalization or Prolonged Hospitalization	A070101	Failure to Capture		
									A0712	Pacing Problem		
									A1301	Failure to Read Input Signal		
					CT2406	Implantable	E0306	Sepsis	A070101	Failure to Capture		
									A0712	Pacing Problem		
									A1301	Failure to Read Input Signal		
							F08	Hospitalization or Prolonged Hospitalization	A070101	Failure to Capture		
									A0712	Pacing Problem		
									A1301	Failure to Read Input Signal		
					CT2407	Active implantable	E0306	Sepsis	A070101	Failure to Capture		
									A0712	Pacing Problem		
									A1301	Failure to Read Input Signal		
							F08	Hospitalization or Prolonged Hospitalization	A070101	Failure to Capture		
									A0712	Pacing Problem		
									A1301	Failure to Read Input Signal		
					CT257	Implantable cardiac pacemakers	E0306	Sepsis	A070101	Failure to Capture		
									A0712	Pacing Problem		
									A1301	Failure to Read Input Signal		
							F08	Hospitalization or Prolonged Hospitalization	A070101	Failure to Capture		
									A0712	Pacing Problem		
									A1301	Failure to Read Input Signal		
					CT2727	Active implant devices/ancillaries	E0306	Sepsis	A070101	Failure to Capture		
									A0712	Pacing Problem		
									A1301	Failure to Read Input Signal		
							F08	Hospitalization or Prolonged Hospitalization	A070101	Failure to Capture		
									A0712	Pacing Problem		
									A1301	Failure to Read Input Signal		

					CT2750	Implantable cardiac electrical stimulators and associated devices	E0306	Sepsis	A070101	Failure to Capture		
									A0712	Pacing Problem		
									A1301	Failure to Read Input Signal		
							F08	Hospitalization or Prolonged Hospitalization	A070101	Failure to Capture		
									A0712	Pacing Problem		
									A1301	Failure to Read Input Signal		
					CT2866	Implantable cardiac electrical stimulators	E0306	Sepsis	A070101	Failure to Capture		
									A0712	Pacing Problem		
									A1301	Failure to Read Input Signal		
							F08	Hospitalization or Prolonged Hospitalization	A070101	Failure to Capture		
									A0712	Pacing Problem		
									A1301	Failure to Read Input Signal		
					CT2985	Active	E0306	Sepsis	A070101	Failure to Capture		
									A0712	Pacing Problem		
									A1301	Failure to Read Input Signal		
							F08	Hospitalization or Prolonged Hospitalization	A070101	Failure to Capture		
									A0712	Pacing Problem		
									A1301	Failure to Read Input Signal		
					CT3001	Implantable metal	E0306	Sepsis	A070101	Failure to Capture		
									A0712	Pacing Problem		
									A1301	Failure to Read Input Signal		
							F08	Hospitalization or Prolonged Hospitalization	A070101	Failure to Capture		
									A0712	Pacing Problem		
									A1301	Failure to Read Input Signal		
					CT3002	Implantable synthetic polymer	E0306	Sepsis	A070101	Failure to Capture		
									A0712	Pacing Problem		
									A1301	Failure to Read Input Signal		
							F08	Hospitalization or Prolonged Hospitalization	A070101	Failure to Capture		
									A0712	Pacing Problem		
									A1301	Failure to Read Input Signal		
					CT3006	Implantable inorganic material	E0306	Sepsis	A070101	Failure to Capture		
									A0712	Pacing Problem		
									A1301	Failure to Read Input Signal		
							F08	Hospitalization or Prolonged Hospitalization	A070101	Failure to Capture		
									A0712	Pacing Problem		
									A1301	Failure to Read Input Signal		
					CT314	Electrically powered	E0306	Sepsis	A070101	Failure to Capture		
									A0712	Pacing Problem		
									A1301	Failure to Read Input Signal		
							F08	Hospitalization or Prolonged Hospitalization	A070101	Failure to Capture		
									A0712	Pacing Problem		
									A1301	Failure to Read Input Signal		
					CT321	Long-term surgical invasive	E0306	Sepsis	A070101	Failure to Capture		
									A0712	Pacing Problem		
									A1301	Failure to Read Input Signal		
							F08	Hospitalization or Prolonged Hospitalization	A070101	Failure to Capture		
									A0712	Pacing Problem		
									A1301	Failure to Read Input Signal		
					CT449	Generators and associated devices	E0306	Sepsis	A070101	Failure to Capture		
									A0712	Pacing Problem		
									A1301	Failure to Read Input Signal		
							F08	Hospitalization or Prolonged Hospitalization	A070101	Failure to Capture		
									A0712	Pacing Problem		
									A1301	Failure to Read Input Signal		
					CT981	Single use	E0306	Sepsis	A070101	Failure to Capture		
									A0712	Pacing Problem		
									A1301	Failure to Read Input Signal		
							F08	Hospitalization or Prolonged Hospitalization	A070101	Failure to Capture		
									A0712	Pacing Problem		
									A1301	Failure to Read Input Signal		
					CT983	Surgical invasive	E0306	Sepsis	A070101	Failure to Capture		
									A0712	Pacing Problem		
									A1301	Failure to Read Input Signal		
							F08	Hospitalization or Prolonged Hospitalization	A070101	Failure to Capture		
									A0712	Pacing Problem		
									A1301	Failure to Read Input Signal		
13358	2023	All other reportable incidents	36241	Pacemaker/defibrillator lead	CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A072201	High Impedance		
							F26	No Health Consequences or Impact	A072201	High Impedance		
13359	2023	All other reportable incidents	47265	Dual-chamber implantable pacemaker, rate-responsive	CT1685	Cardiac pacemakers	E2103	Device Overstimulation of Tissue	A0908	Incorrect, Inadequate or Imprecise Result or Readings	30/05/2023	30/05/2023
							F08	Hospitalization or Prolonged Hospitalization	A0908	Incorrect, Inadequate or Imprecise Result or Readings	30/05/2023	30/05/2023
							F1905	Device Revision or Replacement	A0908	Incorrect, Inadequate or Imprecise Result or Readings	30/05/2023	30/05/2023
					CT257	Implantable cardiac pacemakers	E2103	Device Overstimulation of Tissue	A0908	Incorrect, Inadequate or Imprecise Result or Readings	30/05/2023	30/05/2023
							F08	Hospitalization or Prolonged Hospitalization	A0908	Incorrect, Inadequate or Imprecise Result or Readings	30/05/2023	30/05/2023
							F1905	Device Revision or Replacement	A0908	Incorrect, Inadequate or Imprecise Result or Readings	30/05/2023	30/05/2023
					CT2866	Implantable cardiac electrical stimulators	E2103	Device Overstimulation of Tissue	A0908	Incorrect, Inadequate or Imprecise Result or Readings	30/05/2023	30/05/2023
							F08	Hospitalization or Prolonged Hospitalization	A0908	Incorrect, Inadequate or Imprecise Result or Readings	30/05/2023	30/05/2023
							F1905	Device Revision or Replacement	A0908	Incorrect, Inadequate or Imprecise Result or Readings	30/05/2023	30/05/2023
13360	2023	All other reportable incidents	47267	Single-chamber implantable pacemaker, rate-responsive	CT141	Electrodes/Leads and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	07/06/2023	07/06/2023
							F1901	Additional Surgery	A070504	Premature Discharge of Battery	07/06/2023	07/06/2023
					CT145	Cardiovascular devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	07/06/2023	07/06/2023
							F1901	Additional Surgery	A070504	Premature Discharge of Battery	07/06/2023	07/06/2023

					CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	07/06/2023	07/06/2023
							F1901	Additional Surgery	A070504	Premature Discharge of Battery	07/06/2023	07/06/2023
					CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	07/06/2023	07/06/2023
							F1901	Additional Surgery	A070504	Premature Discharge of Battery	07/06/2023	07/06/2023
					CT233	Surgical	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	07/06/2023	07/06/2023
							F1901	Additional Surgery	A070504	Premature Discharge of Battery	07/06/2023	07/06/2023
					CT2406	Implantable	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	07/06/2023	07/06/2023
							F1901	Additional Surgery	A070504	Premature Discharge of Battery	07/06/2023	07/06/2023
					CT2407	Active implantable	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	07/06/2023	07/06/2023
							F1901	Additional Surgery	A070504	Premature Discharge of Battery	07/06/2023	07/06/2023
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	07/06/2023	07/06/2023
							F1901	Additional Surgery	A070504	Premature Discharge of Battery	07/06/2023	07/06/2023
					CT2727	Active implant devices/ancillaries	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	07/06/2023	07/06/2023
							F1901	Additional Surgery	A070504	Premature Discharge of Battery	07/06/2023	07/06/2023
					CT2750	Implantable cardiac electrical stimulators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	07/06/2023	07/06/2023
							F1901	Additional Surgery	A070504	Premature Discharge of Battery	07/06/2023	07/06/2023
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	07/06/2023	07/06/2023
							F1901	Additional Surgery	A070504	Premature Discharge of Battery	07/06/2023	07/06/2023
					CT2985	Active	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	07/06/2023	07/06/2023
							F1901	Additional Surgery	A070504	Premature Discharge of Battery	07/06/2023	07/06/2023
					CT3001	Implantable metal	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	07/06/2023	07/06/2023
							F1901	Additional Surgery	A070504	Premature Discharge of Battery	07/06/2023	07/06/2023
					CT3002	Implantable synthetic polymer	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	07/06/2023	07/06/2023
							F1901	Additional Surgery	A070504	Premature Discharge of Battery	07/06/2023	07/06/2023
					CT3006	Implantable inorganic material	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	07/06/2023	07/06/2023
							F1901	Additional Surgery	A070504	Premature Discharge of Battery	07/06/2023	07/06/2023
					CT314	Electrically powered	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	07/06/2023	07/06/2023
							F1901	Additional Surgery	A070504	Premature Discharge of Battery	07/06/2023	07/06/2023
					CT321	Long-term surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	07/06/2023	07/06/2023
							F1901	Additional Surgery	A070504	Premature Discharge of Battery	07/06/2023	07/06/2023
					CT449	Generators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	07/06/2023	07/06/2023
							F1901	Additional Surgery	A070504	Premature Discharge of Battery	07/06/2023	07/06/2023
					CT981	Single use	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	07/06/2023	07/06/2023
							F1901	Additional Surgery	A070504	Premature Discharge of Battery	07/06/2023	07/06/2023
					CT983	Surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	07/06/2023	07/06/2023
							F1901	Additional Surgery	A070504	Premature Discharge of Battery	07/06/2023	07/06/2023
13361	2023	All other reportable incidents	47265	Dual-chamber implantable pacemaker, rate-responsive	CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem		
							F1901	Additional Surgery	A24	No Apparent Device Problem		
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem		
							F1901	Additional Surgery	A24	No Apparent Device Problem		
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem		
							F1901	Additional Surgery	A24	No Apparent Device Problem		
13362	2023	All other reportable incidents	47265	Dual-chamber implantable pacemaker, rate-responsive	CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	08/03/2023	08/03/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	08/03/2023	08/03/2023
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	08/03/2023	08/03/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	08/03/2023	08/03/2023
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	08/03/2023	08/03/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	08/03/2023	08/03/2023
13363	2023	All other reportable incidents	47205	Cardiac pulse generator programmer	CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A150105	Key or Button Unresponsive/not Working		
							F26	No Health Consequences or Impact	A150105	Key or Button Unresponsive/not Working		
13364	2023	All other reportable incidents	47265	Dual-chamber implantable pacemaker, rate-responsive	CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
13365	2023	All other reportable incidents	37265	Dual-chamber implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E2104	Electric Shock	A070909	Oversensing		
									A071302	Inappropriate/Inadequate Shock/Stimulation		
									A090801	Signal Artifact/Noise		
									A1301	Failure to Read Input Signal		
							F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A070909	Oversensing		
									A071302	Inappropriate/Inadequate Shock/Stimulation		
									A090801	Signal Artifact/Noise		
									A1301	Failure to Read Input Signal		
13366	2023	All other reportable incidents	17882	Physiologic-monitoring defibrillation system	CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A0719	Unexpected Shutdown		
							F27	Problem Identified During Non-Clinical Procedure	A0719	Unexpected Shutdown		
13367	2023	All other reportable incidents	47265	Dual-chamber implantable pacemaker, rate-responsive	CT141	Electrodes/Leads and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
					CT145	Cardiovascular devices	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
					CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
					CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
					CT233	Surgical	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
					CT2406	Implantable	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		

							F26	No Health Consequences or Impact	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
					CT2407	Active implantable	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
					CT2727	Active implant devices/ancillaries	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
					CT2750	Implantable cardiac electrical stimulators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
					CT2985	Active	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
					CT3001	Implantable metal	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
					CT3002	Implantable synthetic polymer	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
					CT3006	Implantable inorganic material	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
					CT314	Electrically powered	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
					CT321	Long-term surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
					CT449	Generators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
					CT981	Single use	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
					CT983	Surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
13368	2023	All other reportable incidents	35223	Endocardial/interventricular septal pacing lead	CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070909	Oversensing		
							F26	No Health Consequences or Impact	A070909	Oversensing		
13369	2023	All other reportable incidents	47263	Cardiac resynchronization therapy implantable pacemaker	CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A0908	Incorrect, Inadequate or Imprecise Result or Readings	23/11/2022	23/11/2022
							F08	Hospitalization or Prolonged Hospitalization	A0908	Incorrect, Inadequate or Imprecise Result or Readings	23/11/2022	23/11/2022
							F1905	Device Revision or Replacement	A0908	Incorrect, Inadequate or Imprecise Result or Readings	23/11/2022	23/11/2022
					CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A0908	Incorrect, Inadequate or Imprecise Result or Readings	23/11/2022	23/11/2022
							F08	Hospitalization or Prolonged Hospitalization	A0908	Incorrect, Inadequate or Imprecise Result or Readings	23/11/2022	23/11/2022
							F1905	Device Revision or Replacement	A0908	Incorrect, Inadequate or Imprecise Result or Readings	23/11/2022	23/11/2022
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A0908	Incorrect, Inadequate or Imprecise Result or Readings	23/11/2022	23/11/2022
							F08	Hospitalization or Prolonged Hospitalization	A0908	Incorrect, Inadequate or Imprecise Result or Readings	23/11/2022	23/11/2022
							F1905	Device Revision or Replacement	A0908	Incorrect, Inadequate or Imprecise Result or Readings	23/11/2022	23/11/2022
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A0908	Incorrect, Inadequate or Imprecise Result or Readings	23/11/2022	23/11/2022
							F08	Hospitalization or Prolonged Hospitalization	A0908	Incorrect, Inadequate or Imprecise Result or Readings	23/11/2022	23/11/2022
							F1905	Device Revision or Replacement	A0908	Incorrect, Inadequate or Imprecise Result or Readings	23/11/2022	23/11/2022
13370	2023	All other reportable incidents	47205	Cardiac pulse generator programmer	CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A0203	Defective Device		
									A1102	Application Program Problem		
							F26	No Health Consequences or Impact	A0203	Defective Device		
									A1102	Application Program Problem		
13371	2023	All other reportable incidents	17882	Physiologic-monitoring defibrillation system	CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A07	Electrical /Electronic Property Problem		
							F27	Problem Identified During Non-Clinical Procedure	A07	Electrical /Electronic Property Problem		
13372	2023		47265	Dual-chamber implantable pacemaker, rate-	CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	08/03/2023	08/03/2023

				responsive			F1901	Additional Surgery	A24	No Apparent Device Problem	08/03/2023	08/03/2023
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	08/03/2023	08/03/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	08/03/2023	08/03/2023
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	08/03/2023	08/03/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	08/03/2023	08/03/2023
13373	2023	All other reportable incidents	17882	Physiologic-monitoring defibrillation system	CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A0713	Defibrillation/Stimulation Problem		
							F27	Problem Identified During Non-Clinical Procedure	A0713	Defibrillation/Stimulation Problem		
13374	2023	All other reportable incidents	35223	Endocardial/interventricular septal pacing lead	CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070909	Oversensing		
							F12	Serious Injury/ Illness/ Impairment	A070909	Oversensing		
							F23	Unexpected Medical Intervention	A070909	Oversensing		
							F26	No Health Consequences or Impact	A070909	Oversensing		
13375	2023	All other reportable incidents	17882	Physiologic-monitoring defibrillation system	CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A0719	Unexpected Shutdown		
							F27	Problem Identified During Non-Clinical Procedure	A0719	Unexpected Shutdown		
13376	2023	All other reportable incidents	17882	Physiologic-monitoring defibrillation system	CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A090811	Unable to Obtain Readings		
							F26	No Health Consequences or Impact	A090811	Unable to Obtain Readings		
13377	2023	All other reportable incidents	17882	Physiologic-monitoring defibrillation system	CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070803	Failure to Power Up		
							F27	Problem Identified During Non-Clinical Procedure	A070803	Failure to Power Up		
13378	2023	All other reportable incidents	36241	Pacemaker/defibrillator lead	CT1684	Cardiac pacemakers and associated devices	E0604	Cardiac Perforation	A051201	Device Dislodged or Dislocated	13/01/2022	13/01/2022
									A070908	Failure to Sense	13/01/2022	13/01/2022
									A0712	Pacing Problem	13/01/2022	13/01/2022
							F08	Hospitalization or Prolonged Hospitalization	A051201	Device Dislodged or Dislocated	13/01/2022	13/01/2022
									A070908	Failure to Sense	13/01/2022	13/01/2022
									A0712	Pacing Problem	13/01/2022	13/01/2022
							F1203	Life Threatening Illness or Injury	A051201	Device Dislodged or Dislocated	13/01/2022	13/01/2022
									A070908	Failure to Sense	13/01/2022	13/01/2022
									A0712	Pacing Problem	13/01/2022	13/01/2022
							F1903	Device Explantation	A051201	Device Dislodged or Dislocated	13/01/2022	13/01/2022
									A070908	Failure to Sense	13/01/2022	13/01/2022
									A0712	Pacing Problem	13/01/2022	13/01/2022
13379	2023	All other reportable incidents	35223	Endocardial/interventricular septal pacing lead	CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070102	High Capture Threshold	14/06/2023	14/06/2023
							F08	Hospitalization or Prolonged Hospitalization	A070102	High Capture Threshold	14/06/2023	14/06/2023
							F12	Serious Injury/ Illness/ Impairment	A070102	High Capture Threshold	14/06/2023	14/06/2023
							F1905	Device Revision or Replacement	A070102	High Capture Threshold	14/06/2023	14/06/2023
13380	2023	All other reportable incidents	60956	Pacing/defibrillation lead tunneller	CT1684	Cardiac pacemakers and associated devices	E2104	Electric Shock	A070909	Oversensing		
									A071302	Inappropriate/Inadequate Shock/Stimulation		
									A090801	Signal Artifact/Noise		
									A1301	Failure to Read Input Signal		
									A150202	Malposition of Device		
							E210401	Shock from Patient Lead(s)	A070909	Oversensing		
									A071302	Inappropriate/Inadequate Shock/Stimulation		
									A090801	Signal Artifact/Noise		
									A1301	Failure to Read Input Signal		
									A150202	Malposition of Device		
							F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A070909	Oversensing		
									A071302	Inappropriate/Inadequate Shock/Stimulation		
									A090801	Signal Artifact/Noise		
									A1301	Failure to Read Input Signal		
									A150202	Malposition of Device		
							F2203	Imaging Required	A070909	Oversensing		
									A071302	Inappropriate/Inadequate Shock/Stimulation		
									A090801	Signal Artifact/Noise		
									A1301	Failure to Read Input Signal		
									A150202	Malposition of Device		
13381	2023	All other reportable incidents	47270	Cardiac resynchronization therapy implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E210401	Shock from Patient Lead(s)	A070909	Oversensing		
									A071302	Inappropriate/Inadequate Shock/Stimulation		
									A090801	Signal Artifact/Noise		
									A1301	Failure to Read Input Signal		
							F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A070909	Oversensing		
									A071302	Inappropriate/Inadequate Shock/Stimulation		
									A090801	Signal Artifact/Noise		
									A1301	Failure to Read Input Signal		
							F2203	Imaging Required	A070909	Oversensing		
									A071302	Inappropriate/Inadequate Shock/Stimulation		
									A090801	Signal Artifact/Noise		
									A1301	Failure to Read Input Signal		
13382	2023	All other reportable incidents	47263	Cardiac resynchronization therapy implantable pacemaker	CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	23/07/2024	23/07/2024
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	23/07/2024	23/07/2024
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	23/07/2024	23/07/2024
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	23/07/2024	23/07/2024
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	23/07/2024	23/07/2024
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	23/07/2024	23/07/2024
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	23/07/2024	23/07/2024
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	23/07/2024	23/07/2024
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	23/07/2024	23/07/2024
13383	2023	All other reportable incidents	47265	Dual-chamber implantable pacemaker, rate-responsive	CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem	01/11/2023	01/11/2023
									A070504	Premature Discharge of Battery	01/11/2023	01/11/2023
							F08	Hospitalization or Prolonged Hospitalization	A0705	Battery Problem	01/11/2023	01/11/2023
									A070504	Premature Discharge of Battery	01/11/2023	01/11/2023
							F1905	Device Revision or Replacement	A0705	Battery Problem	01/11/2023	01/11/2023
									A070504	Premature Discharge of Battery	01/11/2023	01/11/2023
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem	01/11/2023	01/11/2023
									A070504	Premature Discharge of Battery	01/11/2023	01/11/2023
							F08	Hospitalization or Prolonged Hospitalization	A0705	Battery Problem	01/11/2023	01/11/2023
									A070504	Premature Discharge of Battery	01/11/2023	01/11/2023
							F1905	Device Revision or Replacement	A0705	Battery Problem	01/11/2023	01/11/2023
									A070504	Premature Discharge of Battery	01/11/2023	01/11/2023
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem	01/11/2023	01/11/2023
									A070504	Premature Discharge of Battery	01/11/2023	01/11/2023

							F08	Hospitalization or Prolonged Hospitalization	A0705	Battery Problem	01/11/2023	01/11/2023
									A070504	Premature Discharge of Battery	01/11/2023	01/11/2023
							F1905	Device Revision or Replacement	A0705	Battery Problem	01/11/2023	01/11/2023
									A070504	Premature Discharge of Battery	01/11/2023	01/11/2023
13384	2023	All other reportable incidents	47267	Single-chamber implantable pacemaker, rate-responsive	CT141	Electrodes/Leads and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	14/09/2022	14/09/2022
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	14/09/2022	14/09/2022
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	14/09/2022	14/09/2022
					CT145	Cardiovascular devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	14/09/2022	14/09/2022
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	14/09/2022	14/09/2022
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	14/09/2022	14/09/2022
					CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	14/09/2022	14/09/2022
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	14/09/2022	14/09/2022
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	14/09/2022	14/09/2022
					CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	14/09/2022	14/09/2022
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	14/09/2022	14/09/2022
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	14/09/2022	14/09/2022
					CT233	Surgical	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	14/09/2022	14/09/2022
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	14/09/2022	14/09/2022
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	14/09/2022	14/09/2022
					CT2406	Implantable	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	14/09/2022	14/09/2022
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	14/09/2022	14/09/2022
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	14/09/2022	14/09/2022
					CT2407	Active implantable	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	14/09/2022	14/09/2022
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	14/09/2022	14/09/2022
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	14/09/2022	14/09/2022
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	14/09/2022	14/09/2022
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	14/09/2022	14/09/2022
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	14/09/2022	14/09/2022
					CT2727	Active implant devices/ancillaries	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	14/09/2022	14/09/2022
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	14/09/2022	14/09/2022
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	14/09/2022	14/09/2022
					CT2750	Implantable cardiac electrical stimulators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	14/09/2022	14/09/2022
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	14/09/2022	14/09/2022
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	14/09/2022	14/09/2022
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	14/09/2022	14/09/2022
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	14/09/2022	14/09/2022
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	14/09/2022	14/09/2022
					CT2985	Active	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	14/09/2022	14/09/2022
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	14/09/2022	14/09/2022
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	14/09/2022	14/09/2022
					CT3001	Implantable metal	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	14/09/2022	14/09/2022
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	14/09/2022	14/09/2022
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	14/09/2022	14/09/2022
					CT3002	Implantable synthetic polymer	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	14/09/2022	14/09/2022
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	14/09/2022	14/09/2022
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	14/09/2022	14/09/2022
					CT3006	Implantable inorganic material	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	14/09/2022	14/09/2022
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	14/09/2022	14/09/2022
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	14/09/2022	14/09/2022
					CT314	Electrically powered	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	14/09/2022	14/09/2022
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	14/09/2022	14/09/2022
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	14/09/2022	14/09/2022
					CT321	Long-term surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	14/09/2022	14/09/2022
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	14/09/2022	14/09/2022
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	14/09/2022	14/09/2022
					CT449	Generators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	14/09/2022	14/09/2022
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	14/09/2022	14/09/2022
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	14/09/2022	14/09/2022
					CT981	Single use	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	14/09/2022	14/09/2022
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	14/09/2022	14/09/2022
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	14/09/2022	14/09/2022
					CT983	Surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	14/09/2022	14/09/2022
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	14/09/2022	14/09/2022
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	14/09/2022	14/09/2022
13385	2023	All other reportable incidents	35223	Endocardial/interventricular septal pacing lead	CT1684	Cardiac pacemakers and associated devices	E0112	Dizziness	A0401	Break		
									A070909	Oversensing		
							F26	No Health Consequences or Impact	A0401	Break		
									A070909	Oversensing		
13386	2023	All other reportable incidents	35852	Single-chamber implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
13387	2023	All other reportable incidents	17882	Physiologic-monitoring defibrillation system	CT1684	Cardiac pacemakers and associated devices	E060110	Ventricular Fibrillation	A0713	Defibrillation/Stimulation Problem		
							F26	No Health Consequences or Impact	A0713	Defibrillation/Stimulation Problem		
13388	2023	All other reportable incidents	47267	Single-chamber implantable pacemaker, rate-responsive	CT141	Electrodes/Leads and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A072201	High Impedance		
							F2203	Imaging Required	A072201	High Impedance		
					CT145	Cardiovascular devices	E2403	No Clinical Signs, Symptoms or Conditions	A072201	High Impedance		
							F2203	Imaging Required	A072201	High Impedance		
					CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A072201	High Impedance		
							F2203	Imaging Required	A072201	High Impedance		
					CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A072201	High Impedance		
							F2203	Imaging Required	A072201	High Impedance		
					CT233	Surgical	E2403	No Clinical Signs, Symptoms or Conditions	A072201	High Impedance		
							F2203	Imaging Required	A072201	High Impedance		
					CT2406	Implantable	E2403	No Clinical Signs, Symptoms or Conditions	A072201	High Impedance		
							F2203	Imaging Required	A072201	High Impedance		
					CT2407	Active implantable	E2403	No Clinical Signs, Symptoms or Conditions	A072201	High Impedance		
							F2203	Imaging Required	A072201	High Impedance		

					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A072201	High Impedance		
							F2203	Imaging Required	A072201	High Impedance		
					CT2727	Active implant devices/ancillaries	E2403	No Clinical Signs, Symptoms or Conditions	A072201	High Impedance		
							F2203	Imaging Required	A072201	High Impedance		
					CT2750	Implantable cardiac electrical stimulators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A072201	High Impedance		
							F2203	Imaging Required	A072201	High Impedance		
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A072201	High Impedance		
							F2203	Imaging Required	A072201	High Impedance		
					CT2985	Active	E2403	No Clinical Signs, Symptoms or Conditions	A072201	High Impedance		
							F2203	Imaging Required	A072201	High Impedance		
					CT3001	Implantable metal	E2403	No Clinical Signs, Symptoms or Conditions	A072201	High Impedance		
							F2203	Imaging Required	A072201	High Impedance		
					CT3002	Implantable synthetic polymer	E2403	No Clinical Signs, Symptoms or Conditions	A072201	High Impedance		
							F2203	Imaging Required	A072201	High Impedance		
					CT3006	Implantable inorganic material	E2403	No Clinical Signs, Symptoms or Conditions	A072201	High Impedance		
							F2203	Imaging Required	A072201	High Impedance		
					CT314	Electrically powered	E2403	No Clinical Signs, Symptoms or Conditions	A072201	High Impedance		
							F2203	Imaging Required	A072201	High Impedance		
					CT321	Long-term surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A072201	High Impedance		
							F2203	Imaging Required	A072201	High Impedance		
					CT449	Generators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A072201	High Impedance		
							F2203	Imaging Required	A072201	High Impedance		
					CT981	Single use	E2403	No Clinical Signs, Symptoms or Conditions	A072201	High Impedance		
							F2203	Imaging Required	A072201	High Impedance		
					CT983	Surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A072201	High Impedance		
							F2203	Imaging Required	A072201	High Impedance		
13389	2023	All other reportable incidents	17882	Physiologic-monitoring defibrillation system	CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070912	Incorrect Interpretation of Signal		
							F26	No Health Consequences or Impact	A070912	Incorrect Interpretation of Signal		
13390	2023	All other reportable incidents	37265	Dual-chamber implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E210401	Shock from Patient Lead(s)	A070909	Oversensing		
							F26	No Health Consequences or Impact	A070909	Oversensing		
13391	2023	All other reportable incidents	47265	Dual-chamber implantable pacemaker, rate-responsive	CT141	Electrodes/Leads and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A0908	Incorrect, Inadequate or Imprecise Result or Readings	25/05/2023	25/05/2023
							F08	Hospitalization or Prolonged Hospitalization	A0908	Incorrect, Inadequate or Imprecise Result or Readings	25/05/2023	25/05/2023
							F1905	Device Revision or Replacement	A0908	Incorrect, Inadequate or Imprecise Result or Readings	25/05/2023	25/05/2023
					CT145	Cardiovascular devices	E2403	No Clinical Signs, Symptoms or Conditions	A0908	Incorrect, Inadequate or Imprecise Result or Readings	25/05/2023	25/05/2023
							F08	Hospitalization or Prolonged Hospitalization	A0908	Incorrect, Inadequate or Imprecise Result or Readings	25/05/2023	25/05/2023
							F1905	Device Revision or Replacement	A0908	Incorrect, Inadequate or Imprecise Result or Readings	25/05/2023	25/05/2023
					CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A0908	Incorrect, Inadequate or Imprecise Result or Readings	25/05/2023	25/05/2023
							F08	Hospitalization or Prolonged Hospitalization	A0908	Incorrect, Inadequate or Imprecise Result or Readings	25/05/2023	25/05/2023
							F1905	Device Revision or Replacement	A0908	Incorrect, Inadequate or Imprecise Result or Readings	25/05/2023	25/05/2023
					CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A0908	Incorrect, Inadequate or Imprecise Result or Readings	25/05/2023	25/05/2023
							F08	Hospitalization or Prolonged Hospitalization	A0908	Incorrect, Inadequate or Imprecise Result or Readings	25/05/2023	25/05/2023
							F1905	Device Revision or Replacement	A0908	Incorrect, Inadequate or Imprecise Result or Readings	25/05/2023	25/05/2023
					CT233	Surgical	E2403	No Clinical Signs, Symptoms or Conditions	A0908	Incorrect, Inadequate or Imprecise Result or Readings	25/05/2023	25/05/2023
							F08	Hospitalization or Prolonged Hospitalization	A0908	Incorrect, Inadequate or Imprecise Result or Readings	25/05/2023	25/05/2023
							F1905	Device Revision or Replacement	A0908	Incorrect, Inadequate or Imprecise Result or Readings	25/05/2023	25/05/2023
					CT2406	Implantable	E2403	No Clinical Signs, Symptoms or Conditions	A0908	Incorrect, Inadequate or Imprecise Result or Readings	25/05/2023	25/05/2023
							F08	Hospitalization or Prolonged Hospitalization	A0908	Incorrect, Inadequate or Imprecise Result or Readings	25/05/2023	25/05/2023
							F1905	Device Revision or Replacement	A0908	Incorrect, Inadequate or Imprecise Result or Readings	25/05/2023	25/05/2023
					CT2407	Active implantable	E2403	No Clinical Signs, Symptoms or Conditions	A0908	Incorrect, Inadequate or Imprecise Result or Readings	25/05/2023	25/05/2023
							F08	Hospitalization or Prolonged Hospitalization	A0908	Incorrect, Inadequate or Imprecise Result or Readings	25/05/2023	25/05/2023
							F1905	Device Revision or Replacement	A0908	Incorrect, Inadequate or Imprecise Result or Readings	25/05/2023	25/05/2023
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A0908	Incorrect, Inadequate or Imprecise Result or Readings	25/05/2023	25/05/2023
							F08	Hospitalization or Prolonged Hospitalization	A0908	Incorrect, Inadequate or Imprecise Result or Readings	25/05/2023	25/05/2023
							F1905	Device Revision or Replacement	A0908	Incorrect, Inadequate or Imprecise Result or Readings	25/05/2023	25/05/2023
					CT2727	Active implant devices/ancillaries	E2403	No Clinical Signs, Symptoms or Conditions	A0908	Incorrect, Inadequate or Imprecise Result or Readings	25/05/2023	25/05/2023
							F08	Hospitalization or Prolonged Hospitalization	A0908	Incorrect, Inadequate or Imprecise Result or Readings	25/05/2023	25/05/2023
							F1905	Device Revision or Replacement	A0908	Incorrect, Inadequate or Imprecise Result or Readings	25/05/2023	25/05/2023
					CT2750	Implantable cardiac electrical stimulators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A0908	Incorrect, Inadequate or Imprecise Result or Readings	25/05/2023	25/05/2023
							F08	Hospitalization or Prolonged Hospitalization	A0908	Incorrect, Inadequate or Imprecise Result or Readings	25/05/2023	25/05/2023
							F1905	Device Revision or Replacement	A0908	Incorrect, Inadequate or Imprecise Result or Readings	25/05/2023	25/05/2023
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A0908	Incorrect, Inadequate or Imprecise Result or Readings	25/05/2023	25/05/2023
							F08	Hospitalization or Prolonged Hospitalization	A0908	Incorrect, Inadequate or Imprecise Result or Readings	25/05/2023	25/05/2023
							F1905	Device Revision or Replacement	A0908	Incorrect, Inadequate or Imprecise Result or Readings	25/05/2023	25/05/2023

					CT2985	Active	E2403	No Clinical Signs, Symptoms or Conditions	A0908	Incorrect, Inadequate or Imprecise Result or Readings	25/05/2023	25/05/2023
							F08	Hospitalization or Prolonged Hospitalization	A0908	Incorrect, Inadequate or Imprecise Result or Readings	25/05/2023	25/05/2023
							F1905	Device Revision or Replacement	A0908	Incorrect, Inadequate or Imprecise Result or Readings	25/05/2023	25/05/2023
					CT3001	Implantable metal	E2403	No Clinical Signs, Symptoms or Conditions	A0908	Incorrect, Inadequate or Imprecise Result or Readings	25/05/2023	25/05/2023
							F08	Hospitalization or Prolonged Hospitalization	A0908	Incorrect, Inadequate or Imprecise Result or Readings	25/05/2023	25/05/2023
							F1905	Device Revision or Replacement	A0908	Incorrect, Inadequate or Imprecise Result or Readings	25/05/2023	25/05/2023
					CT3002	Implantable synthetic polymer	E2403	No Clinical Signs, Symptoms or Conditions	A0908	Incorrect, Inadequate or Imprecise Result or Readings	25/05/2023	25/05/2023
							F08	Hospitalization or Prolonged Hospitalization	A0908	Incorrect, Inadequate or Imprecise Result or Readings	25/05/2023	25/05/2023
							F1905	Device Revision or Replacement	A0908	Incorrect, Inadequate or Imprecise Result or Readings	25/05/2023	25/05/2023
					CT3006	Implantable inorganic material	E2403	No Clinical Signs, Symptoms or Conditions	A0908	Incorrect, Inadequate or Imprecise Result or Readings	25/05/2023	25/05/2023
							F08	Hospitalization or Prolonged Hospitalization	A0908	Incorrect, Inadequate or Imprecise Result or Readings	25/05/2023	25/05/2023
							F1905	Device Revision or Replacement	A0908	Incorrect, Inadequate or Imprecise Result or Readings	25/05/2023	25/05/2023
					CT314	Electrically powered	E2403	No Clinical Signs, Symptoms or Conditions	A0908	Incorrect, Inadequate or Imprecise Result or Readings	25/05/2023	25/05/2023
							F08	Hospitalization or Prolonged Hospitalization	A0908	Incorrect, Inadequate or Imprecise Result or Readings	25/05/2023	25/05/2023
							F1905	Device Revision or Replacement	A0908	Incorrect, Inadequate or Imprecise Result or Readings	25/05/2023	25/05/2023
					CT321	Long-term surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A0908	Incorrect, Inadequate or Imprecise Result or Readings	25/05/2023	25/05/2023
							F08	Hospitalization or Prolonged Hospitalization	A0908	Incorrect, Inadequate or Imprecise Result or Readings	25/05/2023	25/05/2023
							F1905	Device Revision or Replacement	A0908	Incorrect, Inadequate or Imprecise Result or Readings	25/05/2023	25/05/2023
					CT449	Generators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A0908	Incorrect, Inadequate or Imprecise Result or Readings	25/05/2023	25/05/2023
							F08	Hospitalization or Prolonged Hospitalization	A0908	Incorrect, Inadequate or Imprecise Result or Readings	25/05/2023	25/05/2023
							F1905	Device Revision or Replacement	A0908	Incorrect, Inadequate or Imprecise Result or Readings	25/05/2023	25/05/2023
					CT981	Single use	E2403	No Clinical Signs, Symptoms or Conditions	A0908	Incorrect, Inadequate or Imprecise Result or Readings	25/05/2023	25/05/2023
							F08	Hospitalization or Prolonged Hospitalization	A0908	Incorrect, Inadequate or Imprecise Result or Readings	25/05/2023	25/05/2023
							F1905	Device Revision or Replacement	A0908	Incorrect, Inadequate or Imprecise Result or Readings	25/05/2023	25/05/2023
					CT983	Surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A0908	Incorrect, Inadequate or Imprecise Result or Readings	25/05/2023	25/05/2023
							F08	Hospitalization or Prolonged Hospitalization	A0908	Incorrect, Inadequate or Imprecise Result or Readings	25/05/2023	25/05/2023
							F1905	Device Revision or Replacement	A0908	Incorrect, Inadequate or Imprecise Result or Readings	25/05/2023	25/05/2023
13392	2023	All other reportable incidents	47267	Single-chamber implantable pacemaker, rate-responsive	CT141	Electrodes/Leads and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070102	High Capture Threshold		
									A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
									A072201	High Impedance		
							F26	No Health Consequences or Impact	A070102	High Capture Threshold		
									A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
									A072201	High Impedance		
					CT145	Cardiovascular devices	E2403	No Clinical Signs, Symptoms or Conditions	A070102	High Capture Threshold		
									A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
									A072201	High Impedance		
							F26	No Health Consequences or Impact	A070102	High Capture Threshold		
									A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
									A072201	High Impedance		
					CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070102	High Capture Threshold		
									A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
									A072201	High Impedance		
							F26	No Health Consequences or Impact	A070102	High Capture Threshold		
									A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
									A072201	High Impedance		
					CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A070102	High Capture Threshold		
									A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
									A072201	High Impedance		
							F26	No Health Consequences or Impact	A070102	High Capture Threshold		
									A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
									A072201	High Impedance		
					CT233	Surgical	E2403	No Clinical Signs, Symptoms or Conditions	A070102	High Capture Threshold		
									A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
									A072201	High Impedance		
							F26	No Health Consequences or Impact	A070102	High Capture Threshold		
									A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
									A072201	High Impedance		

CT2406	Implantable	E2403	No Clinical Signs, Symptoms or Conditions	A070102	High Capture Threshold		
				A0705	Battery Problem		
				A070504	Premature Discharge of Battery		
				A072201	High Impedance		
		F26	No Health Consequences or Impact	A070102	High Capture Threshold		
				A0705	Battery Problem		
				A070504	Premature Discharge of Battery		
				A072201	High Impedance		
CT2407	Active implantable	E2403	No Clinical Signs, Symptoms or Conditions	A070102	High Capture Threshold		
				A0705	Battery Problem		
				A070504	Premature Discharge of Battery		
				A072201	High Impedance		
		F26	No Health Consequences or Impact	A070102	High Capture Threshold		
				A0705	Battery Problem		
				A070504	Premature Discharge of Battery		
				A072201	High Impedance		
CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A070102	High Capture Threshold		
				A0705	Battery Problem		
				A070504	Premature Discharge of Battery		
				A072201	High Impedance		
		F26	No Health Consequences or Impact	A070102	High Capture Threshold		
				A0705	Battery Problem		
				A070504	Premature Discharge of Battery		
				A072201	High Impedance		
CT2727	Active implant devices/ancillaries	E2403	No Clinical Signs, Symptoms or Conditions	A070102	High Capture Threshold		
				A0705	Battery Problem		
				A070504	Premature Discharge of Battery		
				A072201	High Impedance		
		F26	No Health Consequences or Impact	A070102	High Capture Threshold		
				A0705	Battery Problem		
				A070504	Premature Discharge of Battery		
				A072201	High Impedance		
CT2750	Implantable cardiac electrical stimulators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070102	High Capture Threshold		
				A0705	Battery Problem		
				A070504	Premature Discharge of Battery		
				A072201	High Impedance		
		F26	No Health Consequences or Impact	A070102	High Capture Threshold		
				A0705	Battery Problem		
				A070504	Premature Discharge of Battery		
				A072201	High Impedance		
CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A070102	High Capture Threshold		
				A0705	Battery Problem		
				A070504	Premature Discharge of Battery		
				A072201	High Impedance		
		F26	No Health Consequences or Impact	A070102	High Capture Threshold		
				A0705	Battery Problem		
				A070504	Premature Discharge of Battery		
				A072201	High Impedance		
CT2985	Active	E2403	No Clinical Signs, Symptoms or Conditions	A070102	High Capture Threshold		
				A0705	Battery Problem		
				A070504	Premature Discharge of Battery		
				A072201	High Impedance		
		F26	No Health Consequences or Impact	A070102	High Capture Threshold		
				A0705	Battery Problem		
				A070504	Premature Discharge of Battery		
				A072201	High Impedance		
CT3001	Implantable metal	E2403	No Clinical Signs, Symptoms or Conditions	A070102	High Capture Threshold		
				A0705	Battery Problem		
				A070504	Premature Discharge of Battery		
				A072201	High Impedance		
		F26	No Health Consequences or Impact	A070102	High Capture Threshold		
				A0705	Battery Problem		
				A070504	Premature Discharge of Battery		
				A072201	High Impedance		
CT3002	Implantable synthetic polymer	E2403	No Clinical Signs, Symptoms or Conditions	A070102	High Capture Threshold		
				A0705	Battery Problem		
				A070504	Premature Discharge of Battery		
				A072201	High Impedance		
		F26	No Health Consequences or Impact	A070102	High Capture Threshold		
				A0705	Battery Problem		
				A070504	Premature Discharge of Battery		
				A072201	High Impedance		
CT3006	Implantable inorganic material	E2403	No Clinical Signs, Symptoms or Conditions	A070102	High Capture Threshold		
				A0705	Battery Problem		
				A070504	Premature Discharge of Battery		
				A072201	High Impedance		
		F26	No Health Consequences or Impact	A070102	High Capture Threshold		
				A0705	Battery Problem		
				A070504	Premature Discharge of Battery		
				A072201	High Impedance		
CT314	Electrically powered	E2403	No Clinical Signs, Symptoms or Conditions	A070102	High Capture Threshold		
				A0705	Battery Problem		
				A070504	Premature Discharge of Battery		
				A072201	High Impedance		
		F26	No Health Consequences or Impact	A070102	High Capture Threshold		
				A0705	Battery Problem		
				A070504	Premature Discharge of Battery		
				A072201	High Impedance		

					CT321	Long-term surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A070102	High Capture Threshold		
									A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
									A072201	High Impedance		
							F26	No Health Consequences or Impact	A070102	High Capture Threshold		
									A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
									A072201	High Impedance		
					CT449	Generators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070102	High Capture Threshold		
									A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
									A072201	High Impedance		
							F26	No Health Consequences or Impact	A070102	High Capture Threshold		
									A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
									A072201	High Impedance		
					CT981	Single use	E2403	No Clinical Signs, Symptoms or Conditions	A070102	High Capture Threshold		
									A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
									A072201	High Impedance		
							F26	No Health Consequences or Impact	A070102	High Capture Threshold		
									A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
									A072201	High Impedance		
					CT983	Surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A070102	High Capture Threshold		
									A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
									A072201	High Impedance		
							F26	No Health Consequences or Impact	A070102	High Capture Threshold		
									A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
									A072201	High Impedance		
13393	2023	All other reportable incidents	36241	Pacemaker/defibrillator lead	CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070102	High Capture Threshold		
									A072201	High Impedance		
							F26	No Health Consequences or Impact	A070102	High Capture Threshold		
									A072201	High Impedance		
13394	2023	All other reportable incidents	47265	Dual-chamber implantable pacemaker, rate-responsive	CT141	Electrodes/Leads and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A0908	Incorrect, Inadequate or Imprecise Result or Readings		
							F26	No Health Consequences or Impact	A0908	Incorrect, Inadequate or Imprecise Result or Readings		
					CT145	Cardiovascular devices	E2403	No Clinical Signs, Symptoms or Conditions	A0908	Incorrect, Inadequate or Imprecise Result or Readings		
							F26	No Health Consequences or Impact	A0908	Incorrect, Inadequate or Imprecise Result or Readings		
					CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A0908	Incorrect, Inadequate or Imprecise Result or Readings		
							F26	No Health Consequences or Impact	A0908	Incorrect, Inadequate or Imprecise Result or Readings		
					CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A0908	Incorrect, Inadequate or Imprecise Result or Readings		
							F26	No Health Consequences or Impact	A0908	Incorrect, Inadequate or Imprecise Result or Readings		
					CT233	Surgical	E2403	No Clinical Signs, Symptoms or Conditions	A0908	Incorrect, Inadequate or Imprecise Result or Readings		
							F26	No Health Consequences or Impact	A0908	Incorrect, Inadequate or Imprecise Result or Readings		
					CT2406	Implantable	E2403	No Clinical Signs, Symptoms or Conditions	A0908	Incorrect, Inadequate or Imprecise Result or Readings		
							F26	No Health Consequences or Impact	A0908	Incorrect, Inadequate or Imprecise Result or Readings		
					CT2407	Active implantable	E2403	No Clinical Signs, Symptoms or Conditions	A0908	Incorrect, Inadequate or Imprecise Result or Readings		
							F26	No Health Consequences or Impact	A0908	Incorrect, Inadequate or Imprecise Result or Readings		
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A0908	Incorrect, Inadequate or Imprecise Result or Readings		
							F26	No Health Consequences or Impact	A0908	Incorrect, Inadequate or Imprecise Result or Readings		
					CT2727	Active implant devices/ancillaries	E2403	No Clinical Signs, Symptoms or Conditions	A0908	Incorrect, Inadequate or Imprecise Result or Readings		
							F26	No Health Consequences or Impact	A0908	Incorrect, Inadequate or Imprecise Result or Readings		
					CT2750	Implantable cardiac electrical stimulators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A0908	Incorrect, Inadequate or Imprecise Result or Readings		
							F26	No Health Consequences or Impact	A0908	Incorrect, Inadequate or Imprecise Result or Readings		
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A0908	Incorrect, Inadequate or Imprecise Result or Readings		
							F26	No Health Consequences or Impact	A0908	Incorrect, Inadequate or Imprecise Result or Readings		
					CT2985	Active	E2403	No Clinical Signs, Symptoms or Conditions	A0908	Incorrect, Inadequate or Imprecise Result or Readings		
							F26	No Health Consequences or Impact	A0908	Incorrect, Inadequate or Imprecise Result or Readings		
					CT3001	Implantable metal	E2403	No Clinical Signs, Symptoms or Conditions	A0908	Incorrect, Inadequate or Imprecise Result or Readings		
							F26	No Health Consequences or Impact	A0908	Incorrect, Inadequate or Imprecise Result or Readings		
					CT3002	Implantable synthetic polymer	E2403	No Clinical Signs, Symptoms or Conditions	A0908	Incorrect, Inadequate or Imprecise Result or Readings		
							F26	No Health Consequences or Impact	A0908	Incorrect, Inadequate or Imprecise Result or Readings		
					CT3006	Implantable inorganic material	E2403	No Clinical Signs, Symptoms or Conditions	A0908	Incorrect, Inadequate or Imprecise Result or Readings		

						F26	No Health Consequences or Impact	A0908	Incorrect, Inadequate or Imprecise Result or Readings		
					CT314	Electrically powered	E2403	No Clinical Signs, Symptoms or Conditions	A0908	Incorrect, Inadequate or Imprecise Result or Readings		
							F26	No Health Consequences or Impact	A0908	Incorrect, Inadequate or Imprecise Result or Readings		
					CT321	Long-term surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A0908	Incorrect, Inadequate or Imprecise Result or Readings		
							F26	No Health Consequences or Impact	A0908	Incorrect, Inadequate or Imprecise Result or Readings		
					CT449	Generators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A0908	Incorrect, Inadequate or Imprecise Result or Readings		
							F26	No Health Consequences or Impact	A0908	Incorrect, Inadequate or Imprecise Result or Readings		
					CT981	Single use	E2403	No Clinical Signs, Symptoms or Conditions	A0908	Incorrect, Inadequate or Imprecise Result or Readings		
							F26	No Health Consequences or Impact	A0908	Incorrect, Inadequate or Imprecise Result or Readings		
					CT983	Surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A0908	Incorrect, Inadequate or Imprecise Result or Readings		
							F26	No Health Consequences or Impact	A0908	Incorrect, Inadequate or Imprecise Result or Readings		
13395	2023	All other reportable incidents	47267	Single-chamber implantable pacemaker, rate-responsive	CT141	Electrodes/Leads and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	01/07/2023	01/07/2023
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	01/07/2023	01/07/2023
							F1903	Device Explantation	A070504	Premature Discharge of Battery	01/07/2023	01/07/2023
					CT145	Cardiovascular devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	01/07/2023	01/07/2023
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	01/07/2023	01/07/2023
							F1903	Device Explantation	A070504	Premature Discharge of Battery	01/07/2023	01/07/2023
					CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	01/07/2023	01/07/2023
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	01/07/2023	01/07/2023
							F1903	Device Explantation	A070504	Premature Discharge of Battery	01/07/2023	01/07/2023
					CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	01/07/2023	01/07/2023
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	01/07/2023	01/07/2023
							F1903	Device Explantation	A070504	Premature Discharge of Battery	01/07/2023	01/07/2023
					CT233	Surgical	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	01/07/2023	01/07/2023
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	01/07/2023	01/07/2023
							F1903	Device Explantation	A070504	Premature Discharge of Battery	01/07/2023	01/07/2023
					CT2406	Implantable	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	01/07/2023	01/07/2023
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	01/07/2023	01/07/2023
							F1903	Device Explantation	A070504	Premature Discharge of Battery	01/07/2023	01/07/2023
					CT2407	Active implantable	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	01/07/2023	01/07/2023
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	01/07/2023	01/07/2023
							F1903	Device Explantation	A070504	Premature Discharge of Battery	01/07/2023	01/07/2023
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	01/07/2023	01/07/2023
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	01/07/2023	01/07/2023
							F1903	Device Explantation	A070504	Premature Discharge of Battery	01/07/2023	01/07/2023
					CT2727	Active implant devices/ancillaries	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	01/07/2023	01/07/2023
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	01/07/2023	01/07/2023
							F1903	Device Explantation	A070504	Premature Discharge of Battery	01/07/2023	01/07/2023
					CT2750	Implantable cardiac electrical stimulators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	01/07/2023	01/07/2023
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	01/07/2023	01/07/2023
							F1903	Device Explantation	A070504	Premature Discharge of Battery	01/07/2023	01/07/2023
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	01/07/2023	01/07/2023
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	01/07/2023	01/07/2023
							F1903	Device Explantation	A070504	Premature Discharge of Battery	01/07/2023	01/07/2023
					CT2985	Active	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	01/07/2023	01/07/2023
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	01/07/2023	01/07/2023
							F1903	Device Explantation	A070504	Premature Discharge of Battery	01/07/2023	01/07/2023
					CT3001	Implantable metal	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	01/07/2023	01/07/2023
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	01/07/2023	01/07/2023
							F1903	Device Explantation	A070504	Premature Discharge of Battery	01/07/2023	01/07/2023
					CT3002	Implantable synthetic polymer	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	01/07/2023	01/07/2023
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	01/07/2023	01/07/2023
							F1903	Device Explantation	A070504	Premature Discharge of Battery	01/07/2023	01/07/2023
					CT3006	Implantable inorganic material	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	01/07/2023	01/07/2023
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	01/07/2023	01/07/2023
							F1903	Device Explantation	A070504	Premature Discharge of Battery	01/07/2023	01/07/2023
					CT314	Electrically powered	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	01/07/2023	01/07/2023
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	01/07/2023	01/07/2023
							F1903	Device Explantation	A070504	Premature Discharge of Battery	01/07/2023	01/07/2023
					CT321	Long-term surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	01/07/2023	01/07/2023
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	01/07/2023	01/07/2023
							F1903	Device Explantation	A070504	Premature Discharge of Battery	01/07/2023	01/07/2023
					CT449	Generators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	01/07/2023	01/07/2023
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	01/07/2023	01/07/2023
							F1903	Device Explantation	A070504	Premature Discharge of Battery	01/07/2023	01/07/2023
					CT981	Single use	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	01/07/2023	01/07/2023
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	01/07/2023	01/07/2023
							F1903	Device Explantation	A070504	Premature Discharge of Battery	01/07/2023	01/07/2023
					CT983	Surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	01/07/2023	01/07/2023
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	01/07/2023	01/07/2023
							F1903	Device Explantation	A070504	Premature Discharge of Battery	01/07/2023	01/07/2023
13396	2023	All other reportable incidents	47265	Dual-chamber implantable pacemaker, rate-responsive	CT141	Electrodes/Leads and associated devices	E0717	Dyspnea	A071102	Failure to Interrogate	09/06/2023	09/06/2023
									A071204	No Pacing	09/06/2023	09/06/2023
							F1901	Additional Surgery	A071102	Failure to Interrogate	09/06/2023	09/06/2023
									A071204	No Pacing	09/06/2023	09/06/2023
					CT145	Cardiovascular devices	E0717	Dyspnea	A071102	Failure to Interrogate	09/06/2023	09/06/2023
									A071204	No Pacing	09/06/2023	09/06/2023
							F1901	Additional Surgery	A071102	Failure to Interrogate	09/06/2023	09/06/2023
									A071204	No Pacing	09/06/2023	09/06/2023

					CT1684	Cardiac pacemakers and associated devices	E0717	Dyspnea	A071102	Failure to Interrogate	09/06/2023	09/06/2023
									A071204	No Pacing	09/06/2023	09/06/2023
							F1901	Additional Surgery	A071102	Failure to Interrogate	09/06/2023	09/06/2023
									A071204	No Pacing	09/06/2023	09/06/2023
					CT1685	Cardiac pacemakers	E0717	Dyspnea	A071102	Failure to Interrogate	09/06/2023	09/06/2023
									A071204	No Pacing	09/06/2023	09/06/2023
							F1901	Additional Surgery	A071102	Failure to Interrogate	09/06/2023	09/06/2023
									A071204	No Pacing	09/06/2023	09/06/2023
					CT233	Surgical	E0717	Dyspnea	A071102	Failure to Interrogate	09/06/2023	09/06/2023
									A071204	No Pacing	09/06/2023	09/06/2023
							F1901	Additional Surgery	A071102	Failure to Interrogate	09/06/2023	09/06/2023
									A071204	No Pacing	09/06/2023	09/06/2023
					CT2406	Implantable	E0717	Dyspnea	A071102	Failure to Interrogate	09/06/2023	09/06/2023
									A071204	No Pacing	09/06/2023	09/06/2023
							F1901	Additional Surgery	A071102	Failure to Interrogate	09/06/2023	09/06/2023
									A071204	No Pacing	09/06/2023	09/06/2023
					CT2407	Active implantable	E0717	Dyspnea	A071102	Failure to Interrogate	09/06/2023	09/06/2023
									A071204	No Pacing	09/06/2023	09/06/2023
							F1901	Additional Surgery	A071102	Failure to Interrogate	09/06/2023	09/06/2023
									A071204	No Pacing	09/06/2023	09/06/2023
					CT257	Implantable cardiac pacemakers	E0717	Dyspnea	A071102	Failure to Interrogate	09/06/2023	09/06/2023
									A071204	No Pacing	09/06/2023	09/06/2023
							F1901	Additional Surgery	A071102	Failure to Interrogate	09/06/2023	09/06/2023
									A071204	No Pacing	09/06/2023	09/06/2023
					CT2727	Active implant devices/ancillaries	E0717	Dyspnea	A071102	Failure to Interrogate	09/06/2023	09/06/2023
									A071204	No Pacing	09/06/2023	09/06/2023
							F1901	Additional Surgery	A071102	Failure to Interrogate	09/06/2023	09/06/2023
									A071204	No Pacing	09/06/2023	09/06/2023
					CT2750	Implantable cardiac electrical stimulators and associated devices	E0717	Dyspnea	A071102	Failure to Interrogate	09/06/2023	09/06/2023
									A071204	No Pacing	09/06/2023	09/06/2023
							F1901	Additional Surgery	A071102	Failure to Interrogate	09/06/2023	09/06/2023
									A071204	No Pacing	09/06/2023	09/06/2023
					CT2866	Implantable cardiac electrical stimulators	E0717	Dyspnea	A071102	Failure to Interrogate	09/06/2023	09/06/2023
									A071204	No Pacing	09/06/2023	09/06/2023
							F1901	Additional Surgery	A071102	Failure to Interrogate	09/06/2023	09/06/2023
									A071204	No Pacing	09/06/2023	09/06/2023
					CT2985	Active	E0717	Dyspnea	A071102	Failure to Interrogate	09/06/2023	09/06/2023
									A071204	No Pacing	09/06/2023	09/06/2023
							F1901	Additional Surgery	A071102	Failure to Interrogate	09/06/2023	09/06/2023
									A071204	No Pacing	09/06/2023	09/06/2023
					CT3001	Implantable metal	E0717	Dyspnea	A071102	Failure to Interrogate	09/06/2023	09/06/2023
									A071204	No Pacing	09/06/2023	09/06/2023
							F1901	Additional Surgery	A071102	Failure to Interrogate	09/06/2023	09/06/2023
									A071204	No Pacing	09/06/2023	09/06/2023
					CT3002	Implantable synthetic polymer	E0717	Dyspnea	A071102	Failure to Interrogate	09/06/2023	09/06/2023
									A071204	No Pacing	09/06/2023	09/06/2023
							F1901	Additional Surgery	A071102	Failure to Interrogate	09/06/2023	09/06/2023
									A071204	No Pacing	09/06/2023	09/06/2023
					CT3006	Implantable inorganic material	E0717	Dyspnea	A071102	Failure to Interrogate	09/06/2023	09/06/2023
									A071204	No Pacing	09/06/2023	09/06/2023
							F1901	Additional Surgery	A071102	Failure to Interrogate	09/06/2023	09/06/2023
									A071204	No Pacing	09/06/2023	09/06/2023
					CT314	Electrically powered	E0717	Dyspnea	A071102	Failure to Interrogate	09/06/2023	09/06/2023
									A071204	No Pacing	09/06/2023	09/06/2023
							F1901	Additional Surgery	A071102	Failure to Interrogate	09/06/2023	09/06/2023
									A071204	No Pacing	09/06/2023	09/06/2023
					CT321	Long-term surgical invasive	E0717	Dyspnea	A071102	Failure to Interrogate	09/06/2023	09/06/2023
									A071204	No Pacing	09/06/2023	09/06/2023
							F1901	Additional Surgery	A071102	Failure to Interrogate	09/06/2023	09/06/2023
									A071204	No Pacing	09/06/2023	09/06/2023
					CT449	Generators and associated devices	E0717	Dyspnea	A071102	Failure to Interrogate	09/06/2023	09/06/2023
									A071204	No Pacing	09/06/2023	09/06/2023
							F1901	Additional Surgery	A071102	Failure to Interrogate	09/06/2023	09/06/2023
									A071204	No Pacing	09/06/2023	09/06/2023
					CT981	Single use	E0717	Dyspnea	A071102	Failure to Interrogate	09/06/2023	09/06/2023
									A071204	No Pacing	09/06/2023	09/06/2023
							F1901	Additional Surgery	A071102	Failure to Interrogate	09/06/2023	09/06/2023
									A071204	No Pacing	09/06/2023	09/06/2023
					CT983	Surgical invasive	E0717	Dyspnea	A071102	Failure to Interrogate	09/06/2023	09/06/2023
									A071204	No Pacing	09/06/2023	09/06/2023
							F1901	Additional Surgery	A071102	Failure to Interrogate	09/06/2023	09/06/2023
									A071204	No Pacing	09/06/2023	09/06/2023
13397	2023	All other reportable incidents	17882	Physiologic-monitoring defibrillation system	CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A0719	Unexpected Shutdown		
							F26	No Health Consequences or Impact	A0719	Unexpected Shutdown		
13398	2023	All other reportable incidents	47265	Dual-chamber implantable pacemaker, rate-responsive	CT1685	Cardiac pacemakers	E0601	Arrhythmia	A071204	No Pacing	05/04/2023	05/04/2023
									A071302	Inappropriate/Inadequate Shock/Stimulation	05/04/2023	05/04/2023
									A090801	Signal Artifact/Noise	05/04/2023	05/04/2023
							F1901	Additional Surgery	A071204	No Pacing	05/04/2023	05/04/2023
									A071302	Inappropriate/Inadequate Shock/Stimulation	05/04/2023	05/04/2023
									A090801	Signal Artifact/Noise	05/04/2023	05/04/2023
					CT257	Implantable cardiac pacemakers	E0601	Arrhythmia	A071204	No Pacing	05/04/2023	05/04/2023
									A071302	Inappropriate/Inadequate Shock/Stimulation	05/04/2023	05/04/2023
									A090801	Signal Artifact/Noise	05/04/2023	05/04/2023
							F1901	Additional Surgery	A071204	No Pacing	05/04/2023	05/04/2023
									A071302	Inappropriate/Inadequate Shock/Stimulation	05/04/2023	05/04/2023
									A090801	Signal Artifact/Noise	05/04/2023	05/04/2023
					CT2866	Implantable cardiac electrical stimulators	E0601	Arrhythmia	A071204	No Pacing	05/04/2023	05/04/2023
									A071302	Inappropriate/Inadequate Shock/Stimulation	05/04/2023	05/04/2023

									A090801	Signal Artifact/Noise	05/04/2023	05/04/2023
							F1901	Additional Surgery	A071204	No Pacing	05/04/2023	05/04/2023
									A071302	Inappropriate/Inadequate Shock/Stimulation	05/04/2023	05/04/2023
									A090801	Signal Artifact/Noise	05/04/2023	05/04/2023
13399	2023	All other reportable incidents	35852	Single-chamber implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
13400	2023	All other reportable incidents	35852	Single-chamber implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	19/05/2023	19/05/2023
									A0908	Incorrect, Inadequate or Imprecise Result or Readings	19/05/2023	19/05/2023
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery	19/05/2023	19/05/2023
									A0908	Incorrect, Inadequate or Imprecise Result or Readings	19/05/2023	19/05/2023
13401	2023	All other reportable incidents	47265	Dual-chamber implantable pacemaker, rate-responsive	CT141	Electrodes/Leads and associated devices	E0601	Arrhythmia	A070504	Premature Discharge of Battery	11/07/2023	11/07/2023
									A071102	Failure to Interrogate	11/07/2023	11/07/2023
									A071204	No Pacing	11/07/2023	11/07/2023
							E2312	Fatigue	A070504	Premature Discharge of Battery	11/07/2023	11/07/2023
									A071102	Failure to Interrogate	11/07/2023	11/07/2023
									A071204	No Pacing	11/07/2023	11/07/2023
							F1901	Additional Surgery	A070504	Premature Discharge of Battery	11/07/2023	11/07/2023
									A071102	Failure to Interrogate	11/07/2023	11/07/2023
									A071204	No Pacing	11/07/2023	11/07/2023
					CT145	Cardiovascular devices	E0601	Arrhythmia	A070504	Premature Discharge of Battery	11/07/2023	11/07/2023
									A071102	Failure to Interrogate	11/07/2023	11/07/2023
									A071204	No Pacing	11/07/2023	11/07/2023
							E2312	Fatigue	A070504	Premature Discharge of Battery	11/07/2023	11/07/2023
									A071102	Failure to Interrogate	11/07/2023	11/07/2023
									A071204	No Pacing	11/07/2023	11/07/2023
							F1901	Additional Surgery	A070504	Premature Discharge of Battery	11/07/2023	11/07/2023
									A071102	Failure to Interrogate	11/07/2023	11/07/2023
									A071204	No Pacing	11/07/2023	11/07/2023
					CT1684	Cardiac pacemakers and associated devices	E0601	Arrhythmia	A070504	Premature Discharge of Battery	11/07/2023	11/07/2023
									A071102	Failure to Interrogate	11/07/2023	11/07/2023
									A071204	No Pacing	11/07/2023	11/07/2023
							E2312	Fatigue	A070504	Premature Discharge of Battery	11/07/2023	11/07/2023
									A071102	Failure to Interrogate	11/07/2023	11/07/2023
									A071204	No Pacing	11/07/2023	11/07/2023
							F1901	Additional Surgery	A070504	Premature Discharge of Battery	11/07/2023	11/07/2023
									A071102	Failure to Interrogate	11/07/2023	11/07/2023
									A071204	No Pacing	11/07/2023	11/07/2023
					CT1685	Cardiac pacemakers	E0601	Arrhythmia	A070504	Premature Discharge of Battery	11/07/2023	11/07/2023
									A071102	Failure to Interrogate	11/07/2023	11/07/2023
									A071204	No Pacing	11/07/2023	11/07/2023
							E2312	Fatigue	A070504	Premature Discharge of Battery	11/07/2023	11/07/2023
									A071102	Failure to Interrogate	11/07/2023	11/07/2023
									A071204	No Pacing	11/07/2023	11/07/2023
							F1901	Additional Surgery	A070504	Premature Discharge of Battery	11/07/2023	11/07/2023
									A071102	Failure to Interrogate	11/07/2023	11/07/2023
									A071204	No Pacing	11/07/2023	11/07/2023
					CT233	Surgical	E0601	Arrhythmia	A070504	Premature Discharge of Battery	11/07/2023	11/07/2023
									A071102	Failure to Interrogate	11/07/2023	11/07/2023
									A071204	No Pacing	11/07/2023	11/07/2023
							E2312	Fatigue	A070504	Premature Discharge of Battery	11/07/2023	11/07/2023
									A071102	Failure to Interrogate	11/07/2023	11/07/2023
									A071204	No Pacing	11/07/2023	11/07/2023
							F1901	Additional Surgery	A070504	Premature Discharge of Battery	11/07/2023	11/07/2023
									A071102	Failure to Interrogate	11/07/2023	11/07/2023
									A071204	No Pacing	11/07/2023	11/07/2023
					CT2406	Implantable	E0601	Arrhythmia	A070504	Premature Discharge of Battery	11/07/2023	11/07/2023
									A071102	Failure to Interrogate	11/07/2023	11/07/2023
									A071204	No Pacing	11/07/2023	11/07/2023
							E2312	Fatigue	A070504	Premature Discharge of Battery	11/07/2023	11/07/2023
									A071102	Failure to Interrogate	11/07/2023	11/07/2023
									A071204	No Pacing	11/07/2023	11/07/2023
							F1901	Additional Surgery	A070504	Premature Discharge of Battery	11/07/2023	11/07/2023
									A071102	Failure to Interrogate	11/07/2023	11/07/2023
									A071204	No Pacing	11/07/2023	11/07/2023
					CT2407	Active implantable	E0601	Arrhythmia	A070504	Premature Discharge of Battery	11/07/2023	11/07/2023
									A071102	Failure to Interrogate	11/07/2023	11/07/2023
									A071204	No Pacing	11/07/2023	11/07/2023
							E2312	Fatigue	A070504	Premature Discharge of Battery	11/07/2023	11/07/2023
									A071102	Failure to Interrogate	11/07/2023	11/07/2023
									A071204	No Pacing	11/07/2023	11/07/2023
							F1901	Additional Surgery	A070504	Premature Discharge of Battery	11/07/2023	11/07/2023
									A071102	Failure to Interrogate	11/07/2023	11/07/2023
									A071204	No Pacing	11/07/2023	11/07/2023
					CT257	Implantable cardiac pacemakers	E0601	Arrhythmia	A070504	Premature Discharge of Battery	11/07/2023	11/07/2023
									A071102	Failure to Interrogate	11/07/2023	11/07/2023
									A071204	No Pacing	11/07/2023	11/07/2023
							E2312	Fatigue	A070504	Premature Discharge of Battery	11/07/2023	11/07/2023
									A071102	Failure to Interrogate	11/07/2023	11/07/2023
									A071204	No Pacing	11/07/2023	11/07/2023
							F1901	Additional Surgery	A070504	Premature Discharge of Battery	11/07/2023	11/07/2023
									A071102	Failure to Interrogate	11/07/2023	11/07/2023
									A071204	No Pacing	11/07/2023	11/07/2023
					CT2727	Active implant devices/ancillaries	E0601	Arrhythmia	A070504	Premature Discharge of Battery	11/07/2023	11/07/2023
									A071102	Failure to Interrogate	11/07/2023	11/07/2023
									A071204	No Pacing	11/07/2023	11/07/2023
							E2312	Fatigue	A070504	Premature Discharge of Battery	11/07/2023	11/07/2023

								A071102	Failure to Interrogate	11/07/2023	11/07/2023
								A071204	No Pacing	11/07/2023	11/07/2023
							F1901	Additional Surgery	A070504	Premature Discharge of Battery	11/07/2023	11/07/2023
									A071102	Failure to Interrogate	11/07/2023	11/07/2023
									A071204	No Pacing	11/07/2023	11/07/2023
					CT2750	Implantable cardiac electrical stimulators and associated devices	E0601	Arrhythmia	A070504	Premature Discharge of Battery	11/07/2023	11/07/2023
									A071102	Failure to Interrogate	11/07/2023	11/07/2023
									A071204	No Pacing	11/07/2023	11/07/2023
							E2312	Fatigue	A070504	Premature Discharge of Battery	11/07/2023	11/07/2023
									A071102	Failure to Interrogate	11/07/2023	11/07/2023
									A071204	No Pacing	11/07/2023	11/07/2023
							F1901	Additional Surgery	A070504	Premature Discharge of Battery	11/07/2023	11/07/2023
									A071102	Failure to Interrogate	11/07/2023	11/07/2023
									A071204	No Pacing	11/07/2023	11/07/2023
					CT2866	Implantable cardiac electrical stimulators	E0601	Arrhythmia	A070504	Premature Discharge of Battery	11/07/2023	11/07/2023
									A071102	Failure to Interrogate	11/07/2023	11/07/2023
									A071204	No Pacing	11/07/2023	11/07/2023
							E2312	Fatigue	A070504	Premature Discharge of Battery	11/07/2023	11/07/2023
									A071102	Failure to Interrogate	11/07/2023	11/07/2023
									A071204	No Pacing	11/07/2023	11/07/2023
							F1901	Additional Surgery	A070504	Premature Discharge of Battery	11/07/2023	11/07/2023
									A071102	Failure to Interrogate	11/07/2023	11/07/2023
									A071204	No Pacing	11/07/2023	11/07/2023
					CT2985	Active	E0601	Arrhythmia	A070504	Premature Discharge of Battery	11/07/2023	11/07/2023
									A071102	Failure to Interrogate	11/07/2023	11/07/2023
									A071204	No Pacing	11/07/2023	11/07/2023
							E2312	Fatigue	A070504	Premature Discharge of Battery	11/07/2023	11/07/2023
									A071102	Failure to Interrogate	11/07/2023	11/07/2023
									A071204	No Pacing	11/07/2023	11/07/2023
							F1901	Additional Surgery	A070504	Premature Discharge of Battery	11/07/2023	11/07/2023
									A071102	Failure to Interrogate	11/07/2023	11/07/2023
									A071204	No Pacing	11/07/2023	11/07/2023
					CT3001	Implantable metal	E0601	Arrhythmia	A070504	Premature Discharge of Battery	11/07/2023	11/07/2023
									A071102	Failure to Interrogate	11/07/2023	11/07/2023
									A071204	No Pacing	11/07/2023	11/07/2023
							E2312	Fatigue	A070504	Premature Discharge of Battery	11/07/2023	11/07/2023
									A071102	Failure to Interrogate	11/07/2023	11/07/2023
									A071204	No Pacing	11/07/2023	11/07/2023
							F1901	Additional Surgery	A070504	Premature Discharge of Battery	11/07/2023	11/07/2023
									A071102	Failure to Interrogate	11/07/2023	11/07/2023
									A071204	No Pacing	11/07/2023	11/07/2023
					CT3002	Implantable synthetic polymer	E0601	Arrhythmia	A070504	Premature Discharge of Battery	11/07/2023	11/07/2023
									A071102	Failure to Interrogate	11/07/2023	11/07/2023
									A071204	No Pacing	11/07/2023	11/07/2023
							E2312	Fatigue	A070504	Premature Discharge of Battery	11/07/2023	11/07/2023
									A071102	Failure to Interrogate	11/07/2023	11/07/2023
									A071204	No Pacing	11/07/2023	11/07/2023
							F1901	Additional Surgery	A070504	Premature Discharge of Battery	11/07/2023	11/07/2023
									A071102	Failure to Interrogate	11/07/2023	11/07/2023
									A071204	No Pacing	11/07/2023	11/07/2023
					CT3006	Implantable inorganic material	E0601	Arrhythmia	A070504	Premature Discharge of Battery	11/07/2023	11/07/2023
									A071102	Failure to Interrogate	11/07/2023	11/07/2023
									A071204	No Pacing	11/07/2023	11/07/2023
							E2312	Fatigue	A070504	Premature Discharge of Battery	11/07/2023	11/07/2023
									A071102	Failure to Interrogate	11/07/2023	11/07/2023
									A071204	No Pacing	11/07/2023	11/07/2023
							F1901	Additional Surgery	A070504	Premature Discharge of Battery	11/07/2023	11/07/2023
									A071102	Failure to Interrogate	11/07/2023	11/07/2023
									A071204	No Pacing	11/07/2023	11/07/2023
					CT314	Electrically powered	E0601	Arrhythmia	A070504	Premature Discharge of Battery	11/07/2023	11/07/2023
									A071102	Failure to Interrogate	11/07/2023	11/07/2023
									A071204	No Pacing	11/07/2023	11/07/2023
							E2312	Fatigue	A070504	Premature Discharge of Battery	11/07/2023	11/07/2023
									A071102	Failure to Interrogate	11/07/2023	11/07/2023
									A071204	No Pacing	11/07/2023	11/07/2023
							F1901	Additional Surgery	A070504	Premature Discharge of Battery	11/07/2023	11/07/2023
									A071102	Failure to Interrogate	11/07/2023	11/07/2023
									A071204	No Pacing	11/07/2023	11/07/2023
					CT321	Long-term surgical invasive	E0601	Arrhythmia	A070504	Premature Discharge of Battery	11/07/2023	11/07/2023
									A071102	Failure to Interrogate	11/07/2023	11/07/2023
									A071204	No Pacing	11/07/2023	11/07/2023
							E2312	Fatigue	A070504	Premature Discharge of Battery	11/07/2023	11/07/2023
									A071102	Failure to Interrogate	11/07/2023	11/07/2023
									A071204	No Pacing	11/07/2023	11/07/2023
							F1901	Additional Surgery	A070504	Premature Discharge of Battery	11/07/2023	11/07/2023
									A071102	Failure to Interrogate	11/07/2023	11/07/2023
									A071204	No Pacing	11/07/2023	11/07/2023
					CT449	Generators and associated devices	E0601	Arrhythmia	A070504	Premature Discharge of Battery	11/07/2023	11/07/2023
									A071102	Failure to Interrogate	11/07/2023	11/07/2023
									A071204	No Pacing	11/07/2023	11/07/2023
							E2312	Fatigue	A070504	Premature Discharge of Battery	11/07/2023	11/07/2023
									A071102	Failure to Interrogate	11/07/2023	11/07/2023
									A071204	No Pacing	11/07/2023	11/07/2023
							F1901	Additional Surgery	A070504	Premature Discharge of Battery	11/07/2023	11/07/2023
									A071102	Failure to Interrogate	11/07/2023	11/07/2023
									A071204	No Pacing	11/07/2023	11/07/2023
					CT981	Single use	E0601	Arrhythmia	A070504	Premature Discharge of Battery	11/07/2023	11/07/2023
									A071102	Failure to Interrogate	11/07/2023	11/07/2023

									A071204	No Pacing	11/07/2023	11/07/2023
							E2312	Fatigue	A070504	Premature Discharge of Battery	11/07/2023	11/07/2023
									A071102	Failure to Interrogate	11/07/2023	11/07/2023
									A071204	No Pacing	11/07/2023	11/07/2023
							F1901	Additional Surgery	A070504	Premature Discharge of Battery	11/07/2023	11/07/2023
									A071102	Failure to Interrogate	11/07/2023	11/07/2023
									A071204	No Pacing	11/07/2023	11/07/2023
					CT983	Surgical invasive	E0601	Arrhythmia	A070504	Premature Discharge of Battery	11/07/2023	11/07/2023
									A071102	Failure to Interrogate	11/07/2023	11/07/2023
									A071204	No Pacing	11/07/2023	11/07/2023
							E2312	Fatigue	A070504	Premature Discharge of Battery	11/07/2023	11/07/2023
									A071102	Failure to Interrogate	11/07/2023	11/07/2023
									A071204	No Pacing	11/07/2023	11/07/2023
							F1901	Additional Surgery	A070504	Premature Discharge of Battery	11/07/2023	11/07/2023
									A071102	Failure to Interrogate	11/07/2023	11/07/2023
									A071204	No Pacing	11/07/2023	11/07/2023
13402	2023	All other reportable incidents	47263	Cardiac resynchronization therapy implantable pacemaker	CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A0203	Defective Device		
							F26	No Health Consequences or Impact	A0203	Defective Device		
					CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A0203	Defective Device		
							F26	No Health Consequences or Impact	A0203	Defective Device		
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A0203	Defective Device		
							F26	No Health Consequences or Impact	A0203	Defective Device		
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A0203	Defective Device		
							F26	No Health Consequences or Impact	A0203	Defective Device		
13403	2023	All other reportable incidents	47265	Dual-chamber implantable pacemaker, rate-responsive	CT141	Electrodes/Leads and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem	12/01/2023	12/01/2023
									A070504	Premature Discharge of Battery	12/01/2023	12/01/2023
							F08	Hospitalization or Prolonged Hospitalization	A0705	Battery Problem	12/01/2023	12/01/2023
									A070504	Premature Discharge of Battery	12/01/2023	12/01/2023
							F1905	Device Revision or Replacement	A0705	Battery Problem	12/01/2023	12/01/2023
									A070504	Premature Discharge of Battery	12/01/2023	12/01/2023
					CT145	Cardiovascular devices	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem	12/01/2023	12/01/2023
									A070504	Premature Discharge of Battery	12/01/2023	12/01/2023
							F08	Hospitalization or Prolonged Hospitalization	A0705	Battery Problem	12/01/2023	12/01/2023
									A070504	Premature Discharge of Battery	12/01/2023	12/01/2023
							F1905	Device Revision or Replacement	A0705	Battery Problem	12/01/2023	12/01/2023
									A070504	Premature Discharge of Battery	12/01/2023	12/01/2023
					CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem	12/01/2023	12/01/2023
									A070504	Premature Discharge of Battery	12/01/2023	12/01/2023
							F08	Hospitalization or Prolonged Hospitalization	A0705	Battery Problem	12/01/2023	12/01/2023
									A070504	Premature Discharge of Battery	12/01/2023	12/01/2023
							F1905	Device Revision or Replacement	A0705	Battery Problem	12/01/2023	12/01/2023
									A070504	Premature Discharge of Battery	12/01/2023	12/01/2023
					CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem	12/01/2023	12/01/2023
									A070504	Premature Discharge of Battery	12/01/2023	12/01/2023
							F08	Hospitalization or Prolonged Hospitalization	A0705	Battery Problem	12/01/2023	12/01/2023
									A070504	Premature Discharge of Battery	12/01/2023	12/01/2023
							F1905	Device Revision or Replacement	A0705	Battery Problem	12/01/2023	12/01/2023
									A070504	Premature Discharge of Battery	12/01/2023	12/01/2023
					CT233	Surgical	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem	12/01/2023	12/01/2023
									A070504	Premature Discharge of Battery	12/01/2023	12/01/2023
							F08	Hospitalization or Prolonged Hospitalization	A0705	Battery Problem	12/01/2023	12/01/2023
									A070504	Premature Discharge of Battery	12/01/2023	12/01/2023
							F1905	Device Revision or Replacement	A0705	Battery Problem	12/01/2023	12/01/2023
									A070504	Premature Discharge of Battery	12/01/2023	12/01/2023
					CT2406	Implantable	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem	12/01/2023	12/01/2023
									A070504	Premature Discharge of Battery	12/01/2023	12/01/2023
							F08	Hospitalization or Prolonged Hospitalization	A0705	Battery Problem	12/01/2023	12/01/2023
									A070504	Premature Discharge of Battery	12/01/2023	12/01/2023
							F1905	Device Revision or Replacement	A0705	Battery Problem	12/01/2023	12/01/2023
									A070504	Premature Discharge of Battery	12/01/2023	12/01/2023
					CT2407	Active implantable	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem	12/01/2023	12/01/2023
									A070504	Premature Discharge of Battery	12/01/2023	12/01/2023
							F08	Hospitalization or Prolonged Hospitalization	A0705	Battery Problem	12/01/2023	12/01/2023
									A070504	Premature Discharge of Battery	12/01/2023	12/01/2023
							F1905	Device Revision or Replacement	A0705	Battery Problem	12/01/2023	12/01/2023
									A070504	Premature Discharge of Battery	12/01/2023	12/01/2023
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem	12/01/2023	12/01/2023
									A070504	Premature Discharge of Battery	12/01/2023	12/01/2023
							F08	Hospitalization or Prolonged Hospitalization	A0705	Battery Problem	12/01/2023	12/01/2023
									A070504	Premature Discharge of Battery	12/01/2023	12/01/2023
							F1905	Device Revision or Replacement	A0705	Battery Problem	12/01/2023	12/01/2023
									A070504	Premature Discharge of Battery	12/01/2023	12/01/2023
					CT2727	Active implant devices/ancillaries	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem	12/01/2023	12/01/2023
									A070504	Premature Discharge of Battery	12/01/2023	12/01/2023
							F08	Hospitalization or Prolonged Hospitalization	A0705	Battery Problem	12/01/2023	12/01/2023
									A070504	Premature Discharge of Battery	12/01/2023	12/01/2023
							F1905	Device Revision or Replacement	A0705	Battery Problem	12/01/2023	12/01/2023
									A070504	Premature Discharge of Battery	12/01/2023	12/01/2023
					CT2750	Implantable cardiac electrical stimulators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem	12/01/2023	12/01/2023
									A070504	Premature Discharge of Battery	12/01/2023	12/01/2023
							F08	Hospitalization or Prolonged Hospitalization	A0705	Battery Problem	12/01/2023	12/01/2023
									A070504	Premature Discharge of Battery	12/01/2023	12/01/2023
							F1905	Device Revision or Replacement	A0705	Battery Problem	12/01/2023	12/01/2023
									A070504	Premature Discharge of Battery	12/01/2023	12/01/2023
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem	12/01/2023	12/01/2023
									A070504	Premature Discharge of Battery	12/01/2023	12/01/2023
							F08	Hospitalization or Prolonged Hospitalization	A0705	Battery Problem	12/01/2023	12/01/2023
									A070504	Premature Discharge of Battery	12/01/2023	12/01/2023

							F1905	Device Revision or Replacement	A0705	Battery Problem	12/01/2023	12/01/2023
									A070504	Premature Discharge of Battery	12/01/2023	12/01/2023
					CT2985	Active	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem	12/01/2023	12/01/2023
									A070504	Premature Discharge of Battery	12/01/2023	12/01/2023
							F08	Hospitalization or Prolonged Hospitalization	A0705	Battery Problem	12/01/2023	12/01/2023
									A070504	Premature Discharge of Battery	12/01/2023	12/01/2023
							F1905	Device Revision or Replacement	A0705	Battery Problem	12/01/2023	12/01/2023
									A070504	Premature Discharge of Battery	12/01/2023	12/01/2023
					CT3001	Implantable metal	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem	12/01/2023	12/01/2023
									A070504	Premature Discharge of Battery	12/01/2023	12/01/2023
							F08	Hospitalization or Prolonged Hospitalization	A0705	Battery Problem	12/01/2023	12/01/2023
									A070504	Premature Discharge of Battery	12/01/2023	12/01/2023
							F1905	Device Revision or Replacement	A0705	Battery Problem	12/01/2023	12/01/2023
									A070504	Premature Discharge of Battery	12/01/2023	12/01/2023
					CT3002	Implantable synthetic polymer	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem	12/01/2023	12/01/2023
									A070504	Premature Discharge of Battery	12/01/2023	12/01/2023
							F08	Hospitalization or Prolonged Hospitalization	A0705	Battery Problem	12/01/2023	12/01/2023
									A070504	Premature Discharge of Battery	12/01/2023	12/01/2023
							F1905	Device Revision or Replacement	A0705	Battery Problem	12/01/2023	12/01/2023
									A070504	Premature Discharge of Battery	12/01/2023	12/01/2023
					CT3006	Implantable inorganic material	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem	12/01/2023	12/01/2023
									A070504	Premature Discharge of Battery	12/01/2023	12/01/2023
							F08	Hospitalization or Prolonged Hospitalization	A0705	Battery Problem	12/01/2023	12/01/2023
									A070504	Premature Discharge of Battery	12/01/2023	12/01/2023
							F1905	Device Revision or Replacement	A0705	Battery Problem	12/01/2023	12/01/2023
									A070504	Premature Discharge of Battery	12/01/2023	12/01/2023
					CT314	Electrically powered	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem	12/01/2023	12/01/2023
									A070504	Premature Discharge of Battery	12/01/2023	12/01/2023
							F08	Hospitalization or Prolonged Hospitalization	A0705	Battery Problem	12/01/2023	12/01/2023
									A070504	Premature Discharge of Battery	12/01/2023	12/01/2023
							F1905	Device Revision or Replacement	A0705	Battery Problem	12/01/2023	12/01/2023
									A070504	Premature Discharge of Battery	12/01/2023	12/01/2023
					CT321	Long-term surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem	12/01/2023	12/01/2023
									A070504	Premature Discharge of Battery	12/01/2023	12/01/2023
							F08	Hospitalization or Prolonged Hospitalization	A0705	Battery Problem	12/01/2023	12/01/2023
									A070504	Premature Discharge of Battery	12/01/2023	12/01/2023
							F1905	Device Revision or Replacement	A0705	Battery Problem	12/01/2023	12/01/2023
									A070504	Premature Discharge of Battery	12/01/2023	12/01/2023
					CT449	Generators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem	12/01/2023	12/01/2023
									A070504	Premature Discharge of Battery	12/01/2023	12/01/2023
							F08	Hospitalization or Prolonged Hospitalization	A0705	Battery Problem	12/01/2023	12/01/2023
									A070504	Premature Discharge of Battery	12/01/2023	12/01/2023
							F1905	Device Revision or Replacement	A0705	Battery Problem	12/01/2023	12/01/2023
									A070504	Premature Discharge of Battery	12/01/2023	12/01/2023
					CT981	Single use	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem	12/01/2023	12/01/2023
									A070504	Premature Discharge of Battery	12/01/2023	12/01/2023
							F08	Hospitalization or Prolonged Hospitalization	A0705	Battery Problem	12/01/2023	12/01/2023
									A070504	Premature Discharge of Battery	12/01/2023	12/01/2023
							F1905	Device Revision or Replacement	A0705	Battery Problem	12/01/2023	12/01/2023
									A070504	Premature Discharge of Battery	12/01/2023	12/01/2023
					CT983	Surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem	12/01/2023	12/01/2023
									A070504	Premature Discharge of Battery	12/01/2023	12/01/2023
							F08	Hospitalization or Prolonged Hospitalization	A0705	Battery Problem	12/01/2023	12/01/2023
									A070504	Premature Discharge of Battery	12/01/2023	12/01/2023
							F1905	Device Revision or Replacement	A0705	Battery Problem	12/01/2023	12/01/2023
									A070504	Premature Discharge of Battery	12/01/2023	12/01/2023
13404	2023	All other reportable incidents	36241	Pacemaker/defibrillator lead	CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070910	Undersensing		
							F26	No Health Consequences or Impact	A070910	Undersensing		
13405	2023	All other reportable incidents	47265	Dual-chamber implantable pacemaker, rate-responsive	CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem		
							F1901	Additional Surgery	A24	No Apparent Device Problem		
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem		
							F1901	Additional Surgery	A24	No Apparent Device Problem		
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem		
							F1901	Additional Surgery	A24	No Apparent Device Problem		
13406	2023	All other reportable incidents	47205	Cardiac pulse generator programmer	CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A1104	Computer Operating System Problem		
									A13	Communication or Transmission Problem		
							F26	No Health Consequences or Impact	A1104	Computer Operating System Problem		
									A13	Communication or Transmission Problem		
13407	2023	All other reportable incidents	36241	Pacemaker/defibrillator lead	CT1684	Cardiac pacemakers and associated devices	E0604	Cardiac Perforation	A0709	Device Sensing Problem		
									A0712	Pacing Problem		
									A1502	Positioning Problem		
							E0619	Pericardial Effusion	A0709	Device Sensing Problem		
									A0712	Pacing Problem		
									A1502	Positioning Problem		
							E0722	Hemothorax	A0709	Device Sensing Problem		
									A0712	Pacing Problem		
									A1502	Positioning Problem		
							F08	Hospitalization or Prolonged Hospitalization	A0709	Device Sensing Problem		
									A0712	Pacing Problem		
									A1502	Positioning Problem		
							F1203	Life Threatening Illness or Injury	A0709	Device Sensing Problem		
									A0712	Pacing Problem		
									A1502	Positioning Problem		
							F1904	Device Repositioning	A0709	Device Sensing Problem		
									A0712	Pacing Problem		
									A1502	Positioning Problem		
13408	2023	All other reportable incidents	37265	Dual-chamber implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A072201	High Impedance		
									A090801	Signal Artifact/Noise		

							F26	No Health Consequences or Impact	A072201	High Impedance			
									A090801	Signal Artifact/Noise			
13409	2023	All other reportable incidents	47263	Cardiac resynchronization therapy implantable pacemaker	CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070909	Oversensing			
							F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A070909	Oversensing			
					CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A070909	Oversensing			
							F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A070909	Oversensing			
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A070909	Oversensing			
							F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A070909	Oversensing			
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A070909	Oversensing			
							F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A070909	Oversensing			
13410	2023	All other reportable incidents	47265	Dual-chamber implantable pacemaker, rate-responsive	CT1685	Cardiac pacemakers	E060106	Heart Block	A27	Appropriate Device Problem Term/Code not Available			
							F12	Serious Injury/ Illness/ Impairment	A27	Appropriate Device Problem Term/Code not Available			
							F23	Unexpected Medical Intervention	A27	Appropriate Device Problem Term/Code not Available			
					CT257	Implantable cardiac pacemakers	E060106	Heart Block	A27	Appropriate Device Problem Term/Code not Available			
							F12	Serious Injury/ Illness/ Impairment	A27	Appropriate Device Problem Term/Code not Available			
							F23	Unexpected Medical Intervention	A27	Appropriate Device Problem Term/Code not Available			
					CT2866	Implantable cardiac electrical stimulators	E060106	Heart Block	A27	Appropriate Device Problem Term/Code not Available			
							F12	Serious Injury/ Illness/ Impairment	A27	Appropriate Device Problem Term/Code not Available			
							F23	Unexpected Medical Intervention	A27	Appropriate Device Problem Term/Code not Available			
13411	2023	All other reportable incidents	17882	Physiologic-monitoring defibrillation system	CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A0719	Unexpected Shutdown			
							F27	Problem Identified During Non-Clinical Procedure	A0719	Unexpected Shutdown			
13412	2023	All other reportable incidents	36241	Pacemaker/defibrillator lead	CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070909	Oversensing			
									A072201	High Impedance			
							F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A070909	Oversensing			
									A072201	High Impedance			
13413	2023		47265	Dual-chamber implantable pacemaker, rate-responsive	CT141	Electrodes/Leads and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	25/05/2023	25/05/2023	
							F1901	Additional Surgery	A24	No Apparent Device Problem	25/05/2023	25/05/2023	
					CT145	Cardiovascular devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	25/05/2023	25/05/2023	
							F1901	Additional Surgery	A24	No Apparent Device Problem	25/05/2023	25/05/2023	
					CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	25/05/2023	25/05/2023	
							F1901	Additional Surgery	A24	No Apparent Device Problem	25/05/2023	25/05/2023	
					CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	25/05/2023	25/05/2023	
							F1901	Additional Surgery	A24	No Apparent Device Problem	25/05/2023	25/05/2023	
					CT233	Surgical	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	25/05/2023	25/05/2023	
							F1901	Additional Surgery	A24	No Apparent Device Problem	25/05/2023	25/05/2023	
					CT2406	Implantable	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	25/05/2023	25/05/2023	
							F1901	Additional Surgery	A24	No Apparent Device Problem	25/05/2023	25/05/2023	
					CT2407	Active implantable	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	25/05/2023	25/05/2023	
							F1901	Additional Surgery	A24	No Apparent Device Problem	25/05/2023	25/05/2023	
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	25/05/2023	25/05/2023	
							F1901	Additional Surgery	A24	No Apparent Device Problem	25/05/2023	25/05/2023	
					CT2727	Active implant devices/ancillaries	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	25/05/2023	25/05/2023	
							F1901	Additional Surgery	A24	No Apparent Device Problem	25/05/2023	25/05/2023	
					CT2750	Implantable cardiac electrical stimulators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	25/05/2023	25/05/2023	
							F1901	Additional Surgery	A24	No Apparent Device Problem	25/05/2023	25/05/2023	
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	25/05/2023	25/05/2023	
							F1901	Additional Surgery	A24	No Apparent Device Problem	25/05/2023	25/05/2023	
					CT2985	Active	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	25/05/2023	25/05/2023	
							F1901	Additional Surgery	A24	No Apparent Device Problem	25/05/2023	25/05/2023	
					CT3001	Implantable metal	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	25/05/2023	25/05/2023	
							F1901	Additional Surgery	A24	No Apparent Device Problem	25/05/2023	25/05/2023	
					CT3002	Implantable synthetic polymer	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	25/05/2023	25/05/2023	
							F1901	Additional Surgery	A24	No Apparent Device Problem	25/05/2023	25/05/2023	
					CT3006	Implantable inorganic material	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	25/05/2023	25/05/2023	
							F1901	Additional Surgery	A24	No Apparent Device Problem	25/05/2023	25/05/2023	
					CT314	Electrically powered	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	25/05/2023	25/05/2023	
							F1901	Additional Surgery	A24	No Apparent Device Problem	25/05/2023	25/05/2023	
					CT321	Long-term surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	25/05/2023	25/05/2023	
							F1901	Additional Surgery	A24	No Apparent Device Problem	25/05/2023	25/05/2023	
					CT449	Generators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	25/05/2023	25/05/2023	
							F1901	Additional Surgery	A24	No Apparent Device Problem	25/05/2023	25/05/2023	
					CT981	Single use	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	25/05/2023	25/05/2023	
							F1901	Additional Surgery	A24	No Apparent Device Problem	25/05/2023	25/05/2023	
					CT983	Surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	25/05/2023	25/05/2023	
							F1901	Additional Surgery	A24	No Apparent Device Problem	25/05/2023	25/05/2023	
13414	2023	All other reportable incidents	47267	Single-chamber implantable pacemaker, rate-responsive	CT141	Electrodes/Leads and associated devices	E2103	Device Overstimulation of Tissue	A070101	Failure to Capture			
									A070102	High Capture Threshold			
									A071209	Extracardiac Stimulation			
									A1301	Failure to Read Input Signal			
							F26	No Health Consequences or Impact	A070101	Failure to Capture			
									A070102	High Capture Threshold			
									A071209	Extracardiac Stimulation			
									A1301	Failure to Read Input Signal			
					CT145	Cardiovascular devices	E2103	Device Overstimulation of Tissue	A070101	Failure to Capture			
									A070102	High Capture Threshold			
									A071209	Extracardiac Stimulation			
									A1301	Failure to Read Input Signal			
							F26	No Health Consequences or Impact	A070101	Failure to Capture			
									A070102	High Capture Threshold			
									A071209	Extracardiac Stimulation			
									A1301	Failure to Read Input Signal			

CT1684	Cardiac pacemakers and associated devices	E2103	Device Overstimulation of Tissue	A070101	Failure to Capture		
				A070102	High Capture Threshold		
				A071209	Extracardiac Stimulation		
				A1301	Failure to Read Input Signal		
		F26	No Health Consequences or Impact	A070101	Failure to Capture		
				A070102	High Capture Threshold		
				A071209	Extracardiac Stimulation		
				A1301	Failure to Read Input Signal		
CT1685	Cardiac pacemakers	E2103	Device Overstimulation of Tissue	A070101	Failure to Capture		
				A070102	High Capture Threshold		
				A071209	Extracardiac Stimulation		
				A1301	Failure to Read Input Signal		
		F26	No Health Consequences or Impact	A070101	Failure to Capture		
				A070102	High Capture Threshold		
				A071209	Extracardiac Stimulation		
				A1301	Failure to Read Input Signal		
CT233	Surgical	E2103	Device Overstimulation of Tissue	A070101	Failure to Capture		
				A070102	High Capture Threshold		
				A071209	Extracardiac Stimulation		
				A1301	Failure to Read Input Signal		
		F26	No Health Consequences or Impact	A070101	Failure to Capture		
				A070102	High Capture Threshold		
				A071209	Extracardiac Stimulation		
				A1301	Failure to Read Input Signal		
CT2406	Implantable	E2103	Device Overstimulation of Tissue	A070101	Failure to Capture		
				A070102	High Capture Threshold		
				A071209	Extracardiac Stimulation		
				A1301	Failure to Read Input Signal		
		F26	No Health Consequences or Impact	A070101	Failure to Capture		
				A070102	High Capture Threshold		
				A071209	Extracardiac Stimulation		
				A1301	Failure to Read Input Signal		
CT2407	Active implantable	E2103	Device Overstimulation of Tissue	A070101	Failure to Capture		
				A070102	High Capture Threshold		
				A071209	Extracardiac Stimulation		
				A1301	Failure to Read Input Signal		
		F26	No Health Consequences or Impact	A070101	Failure to Capture		
				A070102	High Capture Threshold		
				A071209	Extracardiac Stimulation		
				A1301	Failure to Read Input Signal		
CT257	Implantable cardiac pacemakers	E2103	Device Overstimulation of Tissue	A070101	Failure to Capture		
				A070102	High Capture Threshold		
				A071209	Extracardiac Stimulation		
				A1301	Failure to Read Input Signal		
		F26	No Health Consequences or Impact	A070101	Failure to Capture		
				A070102	High Capture Threshold		
				A071209	Extracardiac Stimulation		
				A1301	Failure to Read Input Signal		
CT2727	Active implant devices/ancillaries	E2103	Device Overstimulation of Tissue	A070101	Failure to Capture		
				A070102	High Capture Threshold		
				A071209	Extracardiac Stimulation		
				A1301	Failure to Read Input Signal		
		F26	No Health Consequences or Impact	A070101	Failure to Capture		
				A070102	High Capture Threshold		
				A071209	Extracardiac Stimulation		
				A1301	Failure to Read Input Signal		
CT2750	Implantable cardiac electrical stimulators and associated devices	E2103	Device Overstimulation of Tissue	A070101	Failure to Capture		
				A070102	High Capture Threshold		
				A071209	Extracardiac Stimulation		
				A1301	Failure to Read Input Signal		
		F26	No Health Consequences or Impact	A070101	Failure to Capture		
				A070102	High Capture Threshold		
				A071209	Extracardiac Stimulation		
				A1301	Failure to Read Input Signal		
CT2866	Implantable cardiac electrical stimulators	E2103	Device Overstimulation of Tissue	A070101	Failure to Capture		
				A070102	High Capture Threshold		
				A071209	Extracardiac Stimulation		
				A1301	Failure to Read Input Signal		
		F26	No Health Consequences or Impact	A070101	Failure to Capture		
				A070102	High Capture Threshold		
				A071209	Extracardiac Stimulation		
				A1301	Failure to Read Input Signal		
CT2985	Active	E2103	Device Overstimulation of Tissue	A070101	Failure to Capture		
				A070102	High Capture Threshold		
				A071209	Extracardiac Stimulation		
				A1301	Failure to Read Input Signal		
		F26	No Health Consequences or Impact	A070101	Failure to Capture		
				A070102	High Capture Threshold		
				A071209	Extracardiac Stimulation		
				A1301	Failure to Read Input Signal		
CT3001	Implantable metal	E2103	Device Overstimulation of Tissue	A070101	Failure to Capture		
				A070102	High Capture Threshold		
				A071209	Extracardiac Stimulation		
				A1301	Failure to Read Input Signal		
		F26	No Health Consequences or Impact	A070101	Failure to Capture		
				A070102	High Capture Threshold		
				A071209	Extracardiac Stimulation		
				A1301	Failure to Read Input Signal		

					CT3002	Implantable synthetic polymer	E2103	Device Overstimulation of Tissue	A070101	Failure to Capture		
									A070102	High Capture Threshold		
									A071209	Extracardiac Stimulation		
									A1301	Failure to Read Input Signal		
							F26	No Health Consequences or Impact	A070101	Failure to Capture		
									A070102	High Capture Threshold		
									A071209	Extracardiac Stimulation		
									A1301	Failure to Read Input Signal		
					CT3006	Implantable inorganic material	E2103	Device Overstimulation of Tissue	A070101	Failure to Capture		
									A070102	High Capture Threshold		
									A071209	Extracardiac Stimulation		
									A1301	Failure to Read Input Signal		
							F26	No Health Consequences or Impact	A070101	Failure to Capture		
									A070102	High Capture Threshold		
									A071209	Extracardiac Stimulation		
									A1301	Failure to Read Input Signal		
					CT314	Electrically powered	E2103	Device Overstimulation of Tissue	A070101	Failure to Capture		
									A070102	High Capture Threshold		
									A071209	Extracardiac Stimulation		
									A1301	Failure to Read Input Signal		
							F26	No Health Consequences or Impact	A070101	Failure to Capture		
									A070102	High Capture Threshold		
									A071209	Extracardiac Stimulation		
									A1301	Failure to Read Input Signal		
					CT321	Long-term surgical invasive	E2103	Device Overstimulation of Tissue	A070101	Failure to Capture		
									A070102	High Capture Threshold		
									A071209	Extracardiac Stimulation		
									A1301	Failure to Read Input Signal		
							F26	No Health Consequences or Impact	A070101	Failure to Capture		
									A070102	High Capture Threshold		
									A071209	Extracardiac Stimulation		
									A1301	Failure to Read Input Signal		
					CT449	Generators and associated devices	E2103	Device Overstimulation of Tissue	A070101	Failure to Capture		
									A070102	High Capture Threshold		
									A071209	Extracardiac Stimulation		
									A1301	Failure to Read Input Signal		
							F26	No Health Consequences or Impact	A070101	Failure to Capture		
									A070102	High Capture Threshold		
									A071209	Extracardiac Stimulation		
									A1301	Failure to Read Input Signal		
					CT981	Single use	E2103	Device Overstimulation of Tissue	A070101	Failure to Capture		
									A070102	High Capture Threshold		
									A071209	Extracardiac Stimulation		
									A1301	Failure to Read Input Signal		
							F26	No Health Consequences or Impact	A070101	Failure to Capture		
									A070102	High Capture Threshold		
									A071209	Extracardiac Stimulation		
									A1301	Failure to Read Input Signal		
					CT983	Surgical invasive	E2103	Device Overstimulation of Tissue	A070101	Failure to Capture		
									A070102	High Capture Threshold		
									A071209	Extracardiac Stimulation		
									A1301	Failure to Read Input Signal		
							F26	No Health Consequences or Impact	A070101	Failure to Capture		
									A070102	High Capture Threshold		
									A071209	Extracardiac Stimulation		
									A1301	Failure to Read Input Signal		
13415	2023	All other reportable incidents	37265	Dual-chamber implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A070909	Oversensing		
							F26	No Health Consequences or Impact	A070909	Oversensing		
13416	2023	All other reportable incidents	35223	Endocardial/interventricular septal pacing lead	CT1684	Cardiac pacemakers and associated devices	E2328	Obstruction/Occlusion	A070102	High Capture Threshold	23/06/2023	23/06/2023
									A0722	Impedance Problem	23/06/2023	23/06/2023
									A072201	High Impedance	23/06/2023	23/06/2023
							F08	Hospitalization or Prolonged Hospitalization	A070102	High Capture Threshold	23/06/2023	23/06/2023
									A0722	Impedance Problem	23/06/2023	23/06/2023
									A072201	High Impedance	23/06/2023	23/06/2023
							F12	Serious Injury/ Illness/ Impairment	A070102	High Capture Threshold	23/06/2023	23/06/2023
									A0722	Impedance Problem	23/06/2023	23/06/2023
									A072201	High Impedance	23/06/2023	23/06/2023
							F23	Unexpected Medical Intervention	A070102	High Capture Threshold	23/06/2023	23/06/2023
									A0722	Impedance Problem	23/06/2023	23/06/2023
									A072201	High Impedance	23/06/2023	23/06/2023
13417	2023	All other reportable incidents	63188	Left-ventricular wireless pacing system sensor/transmitter	CT141	Electrodes/Leads and associated devices			A0708	Power Problem		
					CT145	Cardiovascular devices			A0708	Power Problem		
					CT1577	Wireless monitoring system transmitters			A0708	Power Problem		
					CT1684	Cardiac pacemakers and associated devices			A0708	Power Problem		
					CT2317	Transmitters/Receivers/Transceivers			A0708	Power Problem		
					CT233	Surgical			A0708	Power Problem		
					CT2406	Implantable			A0708	Power Problem		
					CT2407	Active implantable			A0708	Power Problem		
					CT255	Transducers and associated devices			A0708	Power Problem		
					CT2726	Physiological monitoring/measuring devices			A0708	Power Problem		
					CT2727	Active implant devices/ancillaries			A0708	Power Problem		
					CT2750	Implantable cardiac electrical stimulators and associated devices			A0708	Power Problem		
					CT2885	Physiological monitoring/measuring system transmitters/receivers/transceivers			A0708	Power Problem		
					CT2985	Active			A0708	Power Problem		
					CT3001	Implantable metal			A0708	Power Problem		
					CT3002	Implantable synthetic polymer			A0708	Power Problem		

					CT3006	Implantable inorganic material			A0708	Power Problem		
					CT314	Electrically powered			A0708	Power Problem		
					CT321	Long-term surgical invasive			A0708	Power Problem		
					CT384	Sensors and associated devices			A0708	Power Problem		
					CT449	Generators and associated devices			A0708	Power Problem		
					CT981	Single use			A0708	Power Problem		
					CT983	Surgical invasive			A0708	Power Problem		
					CT994	Sensors			A0708	Power Problem		
13418	2023	All other reportable incidents	47205	Cardiac pulse generator programmer	CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A0902	Display or Visual Feedback Problem		
									A11	Computer Software Problem		
							F26	No Health Consequences or Impact	A0902	Display or Visual Feedback Problem		
									A11	Computer Software Problem		
13419	2023	All other reportable incidents	35223	Endocardial/interventricular septal pacing lead	CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070101	Failure to Capture	13/09/2022	13/09/2022
									A070910	Undersensing	13/09/2022	13/09/2022
							F08	Hospitalization or Prolonged Hospitalization	A070101	Failure to Capture	13/09/2022	13/09/2022
									A070910	Undersensing	13/09/2022	13/09/2022
							F12	Serious Injury/ Illness/ Impairment	A070101	Failure to Capture	13/09/2022	13/09/2022
									A070910	Undersensing	13/09/2022	13/09/2022
							F23	Unexpected Medical Intervention	A070101	Failure to Capture	13/09/2022	13/09/2022
									A070910	Undersensing	13/09/2022	13/09/2022
13420	2023	All other reportable incidents	36241	Pacemaker/defibrillator lead	CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A051201	Device Dislodged or Dislocated		
							F26	No Health Consequences or Impact	A051201	Device Dislodged or Dislocated		
13421	2023	All other reportable incidents	36241	Pacemaker/defibrillator lead	CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A072201	High Impedance		
							F26	No Health Consequences or Impact	A072201	High Impedance		
13422	2023	All other reportable incidents	47265	Dual-chamber implantable pacemaker, rate-responsive	CT141	Electrodes/Leads and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070909	Oversensing		
									A072201	High Impedance		
							F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A070909	Oversensing		
									A072201	High Impedance		
					CT145	Cardiovascular devices	E2403	No Clinical Signs, Symptoms or Conditions	A070909	Oversensing		
									A072201	High Impedance		
							F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A070909	Oversensing		
									A072201	High Impedance		
					CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070909	Oversensing		
									A072201	High Impedance		
							F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A070909	Oversensing		
									A072201	High Impedance		
					CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A070909	Oversensing		
									A072201	High Impedance		
							F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A070909	Oversensing		
									A072201	High Impedance		
					CT233	Surgical	E2403	No Clinical Signs, Symptoms or Conditions	A070909	Oversensing		
									A072201	High Impedance		
							F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A070909	Oversensing		
									A072201	High Impedance		
					CT2406	Implantable	E2403	No Clinical Signs, Symptoms or Conditions	A070909	Oversensing		
									A072201	High Impedance		
							F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A070909	Oversensing		
									A072201	High Impedance		
					CT2407	Active implantable	E2403	No Clinical Signs, Symptoms or Conditions	A070909	Oversensing		
									A072201	High Impedance		
							F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A070909	Oversensing		
									A072201	High Impedance		
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A070909	Oversensing		
									A072201	High Impedance		
							F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A070909	Oversensing		
									A072201	High Impedance		
					CT2727	Active implant devices/ancillaries	E2403	No Clinical Signs, Symptoms or Conditions	A070909	Oversensing		
									A072201	High Impedance		
							F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A070909	Oversensing		
									A072201	High Impedance		
					CT2750	Implantable cardiac electrical stimulators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070909	Oversensing		
									A072201	High Impedance		
							F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A070909	Oversensing		
									A072201	High Impedance		
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A070909	Oversensing		
									A072201	High Impedance		
							F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A070909	Oversensing		
									A072201	High Impedance		
					CT2985	Active	E2403	No Clinical Signs, Symptoms or Conditions	A070909	Oversensing		
									A072201	High Impedance		
							F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A070909	Oversensing		
									A072201	High Impedance		
					CT3001	Implantable metal	E2403	No Clinical Signs, Symptoms or Conditions	A070909	Oversensing		
									A072201	High Impedance		
							F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A070909	Oversensing		
									A072201	High Impedance		
					CT3002	Implantable synthetic polymer	E2403	No Clinical Signs, Symptoms or Conditions	A070909	Oversensing		
									A072201	High Impedance		
							F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A070909	Oversensing		
									A072201	High Impedance		
					CT3006	Implantable inorganic material	E2403	No Clinical Signs, Symptoms or Conditions	A070909	Oversensing		
									A072201	High Impedance		
							F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A070909	Oversensing		
									A072201	High Impedance		
					CT314	Electrically powered	E2403	No Clinical Signs, Symptoms or Conditions	A070909	Oversensing		
									A072201	High Impedance		
							F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A070909	Oversensing		
									A072201	High Impedance		

					CT321	Long-term surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A070909	Oversensing		
									A072201	High Impedance		
							F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A070909	Oversensing		
									A072201	High Impedance		
					CT449	Generators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070909	Oversensing		
									A072201	High Impedance		
							F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A070909	Oversensing		
									A072201	High Impedance		
					CT981	Single use	E2403	No Clinical Signs, Symptoms or Conditions	A070909	Oversensing		
									A072201	High Impedance		
							F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A070909	Oversensing		
									A072201	High Impedance		
					CT983	Surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A070909	Oversensing		
									A072201	High Impedance		
							F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A070909	Oversensing		
									A072201	High Impedance		
13423	2023		47265	Dual-chamber implantable pacemaker, rate-responsive	CT141	Electrodes/Leads and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	09/06/2023	09/06/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	09/06/2023	09/06/2023
					CT145	Cardiovascular devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	09/06/2023	09/06/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	09/06/2023	09/06/2023
					CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	09/06/2023	09/06/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	09/06/2023	09/06/2023
					CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	09/06/2023	09/06/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	09/06/2023	09/06/2023
					CT233	Surgical	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	09/06/2023	09/06/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	09/06/2023	09/06/2023
					CT2406	Implantable	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	09/06/2023	09/06/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	09/06/2023	09/06/2023
					CT2407	Active implantable	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	09/06/2023	09/06/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	09/06/2023	09/06/2023
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	09/06/2023	09/06/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	09/06/2023	09/06/2023
					CT2727	Active implant devices/ancillaries	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	09/06/2023	09/06/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	09/06/2023	09/06/2023
					CT2750	Implantable cardiac electrical stimulators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	09/06/2023	09/06/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	09/06/2023	09/06/2023
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	09/06/2023	09/06/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	09/06/2023	09/06/2023
					CT2985	Active	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	09/06/2023	09/06/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	09/06/2023	09/06/2023
					CT3001	Implantable metal	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	09/06/2023	09/06/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	09/06/2023	09/06/2023
					CT3002	Implantable synthetic polymer	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	09/06/2023	09/06/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	09/06/2023	09/06/2023
					CT3006	Implantable inorganic material	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	09/06/2023	09/06/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	09/06/2023	09/06/2023
					CT314	Electrically powered	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	09/06/2023	09/06/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	09/06/2023	09/06/2023
					CT321	Long-term surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	09/06/2023	09/06/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	09/06/2023	09/06/2023
					CT449	Generators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	09/06/2023	09/06/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	09/06/2023	09/06/2023
					CT981	Single use	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	09/06/2023	09/06/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	09/06/2023	09/06/2023
					CT983	Surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	09/06/2023	09/06/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	09/06/2023	09/06/2023
13424	2023	All other reportable incidents	17882	Physiologic-monitoring defibrillation system	CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A150105	Key or Button Unresponsive/not Working		
							F26	No Health Consequences or Impact	A150105	Key or Button Unresponsive/not Working		
13425	2023		47265	Dual-chamber implantable pacemaker, rate-responsive	CT141	Electrodes/Leads and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	08/06/2023	08/06/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	08/06/2023	08/06/2023
					CT145	Cardiovascular devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	08/06/2023	08/06/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	08/06/2023	08/06/2023
					CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	08/06/2023	08/06/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	08/06/2023	08/06/2023
					CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	08/06/2023	08/06/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	08/06/2023	08/06/2023
					CT233	Surgical	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	08/06/2023	08/06/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	08/06/2023	08/06/2023
					CT2406	Implantable	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	08/06/2023	08/06/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	08/06/2023	08/06/2023
					CT2407	Active implantable	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	08/06/2023	08/06/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	08/06/2023	08/06/2023
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	08/06/2023	08/06/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	08/06/2023	08/06/2023
					CT2727	Active implant devices/ancillaries	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	08/06/2023	08/06/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	08/06/2023	08/06/2023
					CT2750	Implantable cardiac electrical stimulators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	08/06/2023	08/06/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	08/06/2023	08/06/2023
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	08/06/2023	08/06/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	08/06/2023	08/06/2023
					CT2985	Active	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	08/06/2023	08/06/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	08/06/2023	08/06/2023
					CT3001	Implantable metal	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	08/06/2023	08/06/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	08/06/2023	08/06/2023
					CT3002	Implantable synthetic polymer	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	08/06/2023	08/06/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	08/06/2023	08/06/2023
					CT3006	Implantable inorganic material	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	08/06/2023	08/06/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	08/06/2023	08/06/2023

					CT314	Electrically powered	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	08/06/2023	08/06/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	08/06/2023	08/06/2023
					CT321	Long-term surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	08/06/2023	08/06/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	08/06/2023	08/06/2023
					CT449	Generators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	08/06/2023	08/06/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	08/06/2023	08/06/2023
					CT981	Single use	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	08/06/2023	08/06/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	08/06/2023	08/06/2023
					CT983	Surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	08/06/2023	08/06/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	08/06/2023	08/06/2023
13426	2023	All other reportable incidents	17882	Physiologic-monitoring defibrillation system	CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A0713	Defibrillation/Stimulation Problem		
							F27	Problem Identified During Non-Clinical Procedure	A0713	Defibrillation/Stimulation Problem		
13427	2023	All other reportable incidents	47265	Dual-chamber implantable pacemaker, rate-responsive	CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem	12/08/2024	12/08/2024
									A070504	Premature Discharge of Battery	12/08/2024	12/08/2024
							F08	Hospitalization or Prolonged Hospitalization	A0705	Battery Problem	12/08/2024	12/08/2024
									A070504	Premature Discharge of Battery	12/08/2024	12/08/2024
							F1905	Device Revision or Replacement	A0705	Battery Problem	12/08/2024	12/08/2024
									A070504	Premature Discharge of Battery	12/08/2024	12/08/2024
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem	12/08/2024	12/08/2024
									A070504	Premature Discharge of Battery	12/08/2024	12/08/2024
							F08	Hospitalization or Prolonged Hospitalization	A0705	Battery Problem	12/08/2024	12/08/2024
									A070504	Premature Discharge of Battery	12/08/2024	12/08/2024
							F1905	Device Revision or Replacement	A0705	Battery Problem	12/08/2024	12/08/2024
									A070504	Premature Discharge of Battery	12/08/2024	12/08/2024
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem	12/08/2024	12/08/2024
									A070504	Premature Discharge of Battery	12/08/2024	12/08/2024
							F08	Hospitalization or Prolonged Hospitalization	A0705	Battery Problem	12/08/2024	12/08/2024
									A070504	Premature Discharge of Battery	12/08/2024	12/08/2024
							F1905	Device Revision or Replacement	A0705	Battery Problem	12/08/2024	12/08/2024
									A070504	Premature Discharge of Battery	12/08/2024	12/08/2024
13428	2023	All other reportable incidents	36241	Pacemaker/defibrillator lead	CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070909	Oversensing		
							F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A070909	Oversensing		
							F2203	Imaging Required	A070909	Oversensing		
13429	2023	All other reportable incidents	36241	Pacemaker/defibrillator lead	CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070910	Undersensing		
									A1301	Failure to Read Input Signal		
							F26	No Health Consequences or Impact	A070910	Undersensing		
									A1301	Failure to Read Input Signal		
13430	2023		63187	Left-ventricular wireless pacing system transducer/electrode	CT1684	Cardiac pacemakers and associated devices	F04	Delay to Diagnosis	A0203	Defective Device		
13431	2023		47265	Dual-chamber implantable pacemaker, rate-responsive	CT141	Electrodes/Leads and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	21/02/2023	21/02/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	21/02/2023	21/02/2023
					CT145	Cardiovascular devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	21/02/2023	21/02/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	21/02/2023	21/02/2023
					CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	21/02/2023	21/02/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	21/02/2023	21/02/2023
					CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	21/02/2023	21/02/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	21/02/2023	21/02/2023
					CT233	Surgical	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	21/02/2023	21/02/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	21/02/2023	21/02/2023
					CT2406	Implantable	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	21/02/2023	21/02/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	21/02/2023	21/02/2023
					CT2407	Active implantable	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	21/02/2023	21/02/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	21/02/2023	21/02/2023
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	21/02/2023	21/02/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	21/02/2023	21/02/2023
					CT2727	Active implant devices/ancillaries	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	21/02/2023	21/02/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	21/02/2023	21/02/2023
					CT2750	Implantable cardiac electrical stimulators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	21/02/2023	21/02/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	21/02/2023	21/02/2023
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	21/02/2023	21/02/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	21/02/2023	21/02/2023
					CT2985	Active	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	21/02/2023	21/02/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	21/02/2023	21/02/2023
					CT3001	Implantable metal	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	21/02/2023	21/02/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	21/02/2023	21/02/2023
					CT3002	Implantable synthetic polymer	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	21/02/2023	21/02/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	21/02/2023	21/02/2023
					CT3006	Implantable inorganic material	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	21/02/2023	21/02/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	21/02/2023	21/02/2023
					CT314	Electrically powered	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	21/02/2023	21/02/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	21/02/2023	21/02/2023
					CT321	Long-term surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	21/02/2023	21/02/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	21/02/2023	21/02/2023
					CT449	Generators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	21/02/2023	21/02/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	21/02/2023	21/02/2023
					CT981	Single use	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	21/02/2023	21/02/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	21/02/2023	21/02/2023
					CT983	Surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	21/02/2023	21/02/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	21/02/2023	21/02/2023
13432	2023	All other reportable incidents	47265	Dual-chamber implantable pacemaker, rate-responsive	CT141	Electrodes/Leads and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A071102	Failure to Interrogate		
									A071204	No Pacing		
							F1901	Additional Surgery	A071102	Failure to Interrogate		
									A071204	No Pacing		
					CT145	Cardiovascular devices	E2403	No Clinical Signs, Symptoms or Conditions	A071102	Failure to Interrogate		
									A071204	No Pacing		
							F1901	Additional Surgery	A071102	Failure to Interrogate		
									A071204	No Pacing		
					CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A071102	Failure to Interrogate		

									A071204	No Pacing		
							F1901	Additional Surgery	A071102	Failure to Interrogate		
									A071204	No Pacing		
					CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A071102	Failure to Interrogate		
									A071204	No Pacing		
							F1901	Additional Surgery	A071102	Failure to Interrogate		
									A071204	No Pacing		
					CT233	Surgical	E2403	No Clinical Signs, Symptoms or Conditions	A071102	Failure to Interrogate		
									A071204	No Pacing		
							F1901	Additional Surgery	A071102	Failure to Interrogate		
									A071204	No Pacing		
					CT2406	Implantable	E2403	No Clinical Signs, Symptoms or Conditions	A071102	Failure to Interrogate		
									A071204	No Pacing		
							F1901	Additional Surgery	A071102	Failure to Interrogate		
									A071204	No Pacing		
					CT2407	Active implantable	E2403	No Clinical Signs, Symptoms or Conditions	A071102	Failure to Interrogate		
									A071204	No Pacing		
							F1901	Additional Surgery	A071102	Failure to Interrogate		
									A071204	No Pacing		
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A071102	Failure to Interrogate		
									A071204	No Pacing		
							F1901	Additional Surgery	A071102	Failure to Interrogate		
									A071204	No Pacing		
					CT2727	Active implant devices/ancillaries	E2403	No Clinical Signs, Symptoms or Conditions	A071102	Failure to Interrogate		
									A071204	No Pacing		
							F1901	Additional Surgery	A071102	Failure to Interrogate		
									A071204	No Pacing		
					CT2750	Implantable cardiac electrical stimulators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A071102	Failure to Interrogate		
									A071204	No Pacing		
							F1901	Additional Surgery	A071102	Failure to Interrogate		
									A071204	No Pacing		
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A071102	Failure to Interrogate		
									A071204	No Pacing		
							F1901	Additional Surgery	A071102	Failure to Interrogate		
									A071204	No Pacing		
					CT2985	Active	E2403	No Clinical Signs, Symptoms or Conditions	A071102	Failure to Interrogate		
									A071204	No Pacing		
							F1901	Additional Surgery	A071102	Failure to Interrogate		
									A071204	No Pacing		
					CT3001	Implantable metal	E2403	No Clinical Signs, Symptoms or Conditions	A071102	Failure to Interrogate		
									A071204	No Pacing		
							F1901	Additional Surgery	A071102	Failure to Interrogate		
									A071204	No Pacing		
					CT3002	Implantable synthetic polymer	E2403	No Clinical Signs, Symptoms or Conditions	A071102	Failure to Interrogate		
									A071204	No Pacing		
							F1901	Additional Surgery	A071102	Failure to Interrogate		
									A071204	No Pacing		
					CT3006	Implantable inorganic material	E2403	No Clinical Signs, Symptoms or Conditions	A071102	Failure to Interrogate		
									A071204	No Pacing		
							F1901	Additional Surgery	A071102	Failure to Interrogate		
									A071204	No Pacing		
					CT314	Electrically powered	E2403	No Clinical Signs, Symptoms or Conditions	A071102	Failure to Interrogate		
									A071204	No Pacing		
							F1901	Additional Surgery	A071102	Failure to Interrogate		
									A071204	No Pacing		
					CT321	Long-term surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A071102	Failure to Interrogate		
									A071204	No Pacing		
							F1901	Additional Surgery	A071102	Failure to Interrogate		
									A071204	No Pacing		
					CT449	Generators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A071102	Failure to Interrogate		
									A071204	No Pacing		
							F1901	Additional Surgery	A071102	Failure to Interrogate		
									A071204	No Pacing		
					CT981	Single use	E2403	No Clinical Signs, Symptoms or Conditions	A071102	Failure to Interrogate		
									A071204	No Pacing		
							F1901	Additional Surgery	A071102	Failure to Interrogate		
									A071204	No Pacing		
					CT983	Surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A071102	Failure to Interrogate		
									A071204	No Pacing		
							F1901	Additional Surgery	A071102	Failure to Interrogate		
									A071204	No Pacing		
13433	2023	All other reportable incidents	47265	Dual-chamber implantable pacemaker, rate-responsive	CT141	Electrodes/Leads and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
							F24	Insufficient Health Impact Information	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
					CT145	Cardiovascular devices	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
							F24	Insufficient Health Impact Information	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
					CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
							F24	Insufficient Health Impact Information	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
					CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
							F24	Insufficient Health Impact Information	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
					CT233	Surgical	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem		

									A070504	Premature Discharge of Battery		
							F24	Insufficient Health Impact Information	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
					CT2406	Implantable	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
							F24	Insufficient Health Impact Information	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
					CT2407	Active implantable	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
							F24	Insufficient Health Impact Information	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
							F24	Insufficient Health Impact Information	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
					CT2727	Active implant devices/ancillaries	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
							F24	Insufficient Health Impact Information	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
					CT2750	Implantable cardiac electrical stimulators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
							F24	Insufficient Health Impact Information	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
							F24	Insufficient Health Impact Information	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
					CT2985	Active	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
							F24	Insufficient Health Impact Information	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
					CT3001	Implantable metal	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
							F24	Insufficient Health Impact Information	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
					CT3002	Implantable synthetic polymer	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
							F24	Insufficient Health Impact Information	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
					CT3006	Implantable inorganic material	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
							F24	Insufficient Health Impact Information	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
					CT314	Electrically powered	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
							F24	Insufficient Health Impact Information	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
					CT321	Long-term surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
							F24	Insufficient Health Impact Information	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
					CT449	Generators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
							F24	Insufficient Health Impact Information	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
					CT981	Single use	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
							F24	Insufficient Health Impact Information	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
					CT983	Surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
							F24	Insufficient Health Impact Information	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
13434	2023	All other reportable incidents	47265	Dual-chamber implantable pacemaker, rate-responsive	CT141	Electrodes/Leads and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	09/06/2023	09/06/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	09/06/2023	09/06/2023
					CT145	Cardiovascular devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	09/06/2023	09/06/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	09/06/2023	09/06/2023
					CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	09/06/2023	09/06/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	09/06/2023	09/06/2023
					CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	09/06/2023	09/06/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	09/06/2023	09/06/2023
					CT233	Surgical	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	09/06/2023	09/06/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	09/06/2023	09/06/2023
					CT2406	Implantable	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	09/06/2023	09/06/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	09/06/2023	09/06/2023
					CT2407	Active implantable	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	09/06/2023	09/06/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	09/06/2023	09/06/2023
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	09/06/2023	09/06/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	09/06/2023	09/06/2023
					CT2727	Active implant devices/ancillaries	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	09/06/2023	09/06/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	09/06/2023	09/06/2023
					CT2750	Implantable cardiac electrical stimulators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	09/06/2023	09/06/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	09/06/2023	09/06/2023
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	09/06/2023	09/06/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	09/06/2023	09/06/2023
					CT2985	Active	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	09/06/2023	09/06/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	09/06/2023	09/06/2023
					CT3001	Implantable metal	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	09/06/2023	09/06/2023

							F1901	Additional Surgery	A24	No Apparent Device Problem	09/06/2023	09/06/2023
					CT3002	Implantable synthetic polymer	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	09/06/2023	09/06/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	09/06/2023	09/06/2023
					CT3006	Implantable inorganic material	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	09/06/2023	09/06/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	09/06/2023	09/06/2023
					CT314	Electrically powered	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	09/06/2023	09/06/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	09/06/2023	09/06/2023
					CT321	Long-term surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	09/06/2023	09/06/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	09/06/2023	09/06/2023
					CT449	Generators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	09/06/2023	09/06/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	09/06/2023	09/06/2023
					CT981	Single use	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	09/06/2023	09/06/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	09/06/2023	09/06/2023
					CT983	Surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	09/06/2023	09/06/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	09/06/2023	09/06/2023
13435	2023		35223	Endocardial/interventricular septal pacing lead	CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A040609	Material Twisted/Bent	15/06/2023	15/06/2023
							F1908	Prolonged Surgery	A040609	Material Twisted/Bent	15/06/2023	15/06/2023
13436	2023		47265	Dual-chamber implantable pacemaker, rate-responsive	CT141	Electrodes/Leads and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A150207	Difficult to Remove	07/06/2023	07/06/2023
									A24	No Apparent Device Problem	07/06/2023	07/06/2023
							F1901	Additional Surgery	A150207	Difficult to Remove	07/06/2023	07/06/2023
									A24	No Apparent Device Problem	07/06/2023	07/06/2023
					CT145	Cardiovascular devices	E2403	No Clinical Signs, Symptoms or Conditions	A150207	Difficult to Remove	07/06/2023	07/06/2023
									A24	No Apparent Device Problem	07/06/2023	07/06/2023
							F1901	Additional Surgery	A150207	Difficult to Remove	07/06/2023	07/06/2023
									A24	No Apparent Device Problem	07/06/2023	07/06/2023
					CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A150207	Difficult to Remove	07/06/2023	07/06/2023
									A24	No Apparent Device Problem	07/06/2023	07/06/2023
							F1901	Additional Surgery	A150207	Difficult to Remove	07/06/2023	07/06/2023
									A24	No Apparent Device Problem	07/06/2023	07/06/2023
					CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A150207	Difficult to Remove	07/06/2023	07/06/2023
									A24	No Apparent Device Problem	07/06/2023	07/06/2023
							F1901	Additional Surgery	A150207	Difficult to Remove	07/06/2023	07/06/2023
									A24	No Apparent Device Problem	07/06/2023	07/06/2023
					CT233	Surgical	E2403	No Clinical Signs, Symptoms or Conditions	A150207	Difficult to Remove	07/06/2023	07/06/2023
									A24	No Apparent Device Problem	07/06/2023	07/06/2023
							F1901	Additional Surgery	A150207	Difficult to Remove	07/06/2023	07/06/2023
									A24	No Apparent Device Problem	07/06/2023	07/06/2023
					CT2406	Implantable	E2403	No Clinical Signs, Symptoms or Conditions	A150207	Difficult to Remove	07/06/2023	07/06/2023
									A24	No Apparent Device Problem	07/06/2023	07/06/2023
							F1901	Additional Surgery	A150207	Difficult to Remove	07/06/2023	07/06/2023
									A24	No Apparent Device Problem	07/06/2023	07/06/2023
					CT2407	Active implantable	E2403	No Clinical Signs, Symptoms or Conditions	A150207	Difficult to Remove	07/06/2023	07/06/2023
									A24	No Apparent Device Problem	07/06/2023	07/06/2023
							F1901	Additional Surgery	A150207	Difficult to Remove	07/06/2023	07/06/2023
									A24	No Apparent Device Problem	07/06/2023	07/06/2023
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A150207	Difficult to Remove	07/06/2023	07/06/2023
									A24	No Apparent Device Problem	07/06/2023	07/06/2023
							F1901	Additional Surgery	A150207	Difficult to Remove	07/06/2023	07/06/2023
									A24	No Apparent Device Problem	07/06/2023	07/06/2023
					CT2727	Active implant devices/ancillaries	E2403	No Clinical Signs, Symptoms or Conditions	A150207	Difficult to Remove	07/06/2023	07/06/2023
									A24	No Apparent Device Problem	07/06/2023	07/06/2023
							F1901	Additional Surgery	A150207	Difficult to Remove	07/06/2023	07/06/2023
									A24	No Apparent Device Problem	07/06/2023	07/06/2023
					CT2750	Implantable cardiac electrical stimulators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A150207	Difficult to Remove	07/06/2023	07/06/2023
									A24	No Apparent Device Problem	07/06/2023	07/06/2023
							F1901	Additional Surgery	A150207	Difficult to Remove	07/06/2023	07/06/2023
									A24	No Apparent Device Problem	07/06/2023	07/06/2023
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A150207	Difficult to Remove	07/06/2023	07/06/2023
									A24	No Apparent Device Problem	07/06/2023	07/06/2023
							F1901	Additional Surgery	A150207	Difficult to Remove	07/06/2023	07/06/2023
									A24	No Apparent Device Problem	07/06/2023	07/06/2023
					CT2985	Active	E2403	No Clinical Signs, Symptoms or Conditions	A150207	Difficult to Remove	07/06/2023	07/06/2023
									A24	No Apparent Device Problem	07/06/2023	07/06/2023
							F1901	Additional Surgery	A150207	Difficult to Remove	07/06/2023	07/06/2023
									A24	No Apparent Device Problem	07/06/2023	07/06/2023
					CT3001	Implantable metal	E2403	No Clinical Signs, Symptoms or Conditions	A150207	Difficult to Remove	07/06/2023	07/06/2023
									A24	No Apparent Device Problem	07/06/2023	07/06/2023
							F1901	Additional Surgery	A150207	Difficult to Remove	07/06/2023	07/06/2023
									A24	No Apparent Device Problem	07/06/2023	07/06/2023
					CT3002	Implantable synthetic polymer	E2403	No Clinical Signs, Symptoms or Conditions	A150207	Difficult to Remove	07/06/2023	07/06/2023
									A24	No Apparent Device Problem	07/06/2023	07/06/2023
							F1901	Additional Surgery	A150207	Difficult to Remove	07/06/2023	07/06/2023
									A24	No Apparent Device Problem	07/06/2023	07/06/2023
					CT3006	Implantable inorganic material	E2403	No Clinical Signs, Symptoms or Conditions	A150207	Difficult to Remove	07/06/2023	07/06/2023
									A24	No Apparent Device Problem	07/06/2023	07/06/2023
							F1901	Additional Surgery	A150207	Difficult to Remove	07/06/2023	07/06/2023
									A24	No Apparent Device Problem	07/06/2023	07/06/2023
					CT314	Electrically powered	E2403	No Clinical Signs, Symptoms or Conditions	A150207	Difficult to Remove	07/06/2023	07/06/2023
									A24	No Apparent Device Problem	07/06/2023	07/06/2023
							F1901	Additional Surgery	A150207	Difficult to Remove	07/06/2023	07/06/2023
									A24	No Apparent Device Problem	07/06/2023	07/06/2023
					CT321	Long-term surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A150207	Difficult to Remove	07/06/2023	07/06/2023
									A24	No Apparent Device Problem	07/06/2023	07/06/2023
							F1901	Additional Surgery	A150207	Difficult to Remove	07/06/2023	07/06/2023
									A24	No Apparent Device Problem	07/06/2023	07/06/2023
					CT449	Generators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A150207	Difficult to Remove	07/06/2023	07/06/2023
									A24	No Apparent Device Problem	07/06/2023	07/06/2023
							F1901	Additional Surgery	A150207	Difficult to Remove	07/06/2023	07/06/2023

								A24	No Apparent Device Problem	07/06/2023	07/06/2023
				CT981	Single use	E2403	No Clinical Signs, Symptoms or Conditions	A150207	Difficult to Remove	07/06/2023	07/06/2023
								A24	No Apparent Device Problem	07/06/2023	07/06/2023
						F1901	Additional Surgery	A150207	Difficult to Remove	07/06/2023	07/06/2023
								A24	No Apparent Device Problem	07/06/2023	07/06/2023
				CT983	Surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A150207	Difficult to Remove	07/06/2023	07/06/2023
								A24	No Apparent Device Problem	07/06/2023	07/06/2023
						F1901	Additional Surgery	A150207	Difficult to Remove	07/06/2023	07/06/2023
								A24	No Apparent Device Problem	07/06/2023	07/06/2023
13437	2023	All other reportable incidents	17882	Physiologic-monitoring defibrillation system	CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A07	Electrical /Electronic Property Problem		
							F27	Problem Identified During Non-Clinical Procedure	A07	Electrical /Electronic Property Problem		
13438	2023	All other reportable incidents	17882	Physiologic-monitoring defibrillation system	CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A0713	Defibrillation/Stimulation Problem		
							F27	Problem Identified During Non-Clinical Procedure	A0713	Defibrillation/Stimulation Problem		
13439	2023	All other reportable incidents	47265	Dual-chamber implantable pacemaker, rate-responsive	CT141	Electrodes/Leads and associated devices	E011903	Syncope/Fainting	A070103	Intermittent Capture	25/01/2023	25/01/2023
							F1905	Device Revision or Replacement	A070103	Intermittent Capture	25/01/2023	25/01/2023
					CT145	Cardiovascular devices	E011903	Syncope/Fainting	A070103	Intermittent Capture	25/01/2023	25/01/2023
							F1905	Device Revision or Replacement	A070103	Intermittent Capture	25/01/2023	25/01/2023
					CT1684	Cardiac pacemakers and associated devices	E011903	Syncope/Fainting	A070103	Intermittent Capture	25/01/2023	25/01/2023
							F1905	Device Revision or Replacement	A070103	Intermittent Capture	25/01/2023	25/01/2023
					CT1685	Cardiac pacemakers	E011903	Syncope/Fainting	A070103	Intermittent Capture	25/01/2023	25/01/2023
							F1905	Device Revision or Replacement	A070103	Intermittent Capture	25/01/2023	25/01/2023
					CT233	Surgical	E011903	Syncope/Fainting	A070103	Intermittent Capture	25/01/2023	25/01/2023
							F1905	Device Revision or Replacement	A070103	Intermittent Capture	25/01/2023	25/01/2023
					CT2406	Implantable	E011903	Syncope/Fainting	A070103	Intermittent Capture	25/01/2023	25/01/2023
							F1905	Device Revision or Replacement	A070103	Intermittent Capture	25/01/2023	25/01/2023
					CT2407	Active implantable	E011903	Syncope/Fainting	A070103	Intermittent Capture	25/01/2023	25/01/2023
							F1905	Device Revision or Replacement	A070103	Intermittent Capture	25/01/2023	25/01/2023
					CT257	Implantable cardiac pacemakers	E011903	Syncope/Fainting	A070103	Intermittent Capture	25/01/2023	25/01/2023
							F1905	Device Revision or Replacement	A070103	Intermittent Capture	25/01/2023	25/01/2023
					CT2727	Active implant devices/ancillaries	E011903	Syncope/Fainting	A070103	Intermittent Capture	25/01/2023	25/01/2023
							F1905	Device Revision or Replacement	A070103	Intermittent Capture	25/01/2023	25/01/2023
					CT2750	Implantable cardiac electrical stimulators and associated devices	E011903	Syncope/Fainting	A070103	Intermittent Capture	25/01/2023	25/01/2023
							F1905	Device Revision or Replacement	A070103	Intermittent Capture	25/01/2023	25/01/2023
					CT2866	Implantable cardiac electrical stimulators	E011903	Syncope/Fainting	A070103	Intermittent Capture	25/01/2023	25/01/2023
							F1905	Device Revision or Replacement	A070103	Intermittent Capture	25/01/2023	25/01/2023
					CT2985	Active	E011903	Syncope/Fainting	A070103	Intermittent Capture	25/01/2023	25/01/2023
							F1905	Device Revision or Replacement	A070103	Intermittent Capture	25/01/2023	25/01/2023
					CT3001	Implantable metal	E011903	Syncope/Fainting	A070103	Intermittent Capture	25/01/2023	25/01/2023
							F1905	Device Revision or Replacement	A070103	Intermittent Capture	25/01/2023	25/01/2023
					CT3002	Implantable synthetic polymer	E011903	Syncope/Fainting	A070103	Intermittent Capture	25/01/2023	25/01/2023
							F1905	Device Revision or Replacement	A070103	Intermittent Capture	25/01/2023	25/01/2023
					CT3006	Implantable inorganic material	E011903	Syncope/Fainting	A070103	Intermittent Capture	25/01/2023	25/01/2023
							F1905	Device Revision or Replacement	A070103	Intermittent Capture	25/01/2023	25/01/2023
					CT314	Electrically powered	E011903	Syncope/Fainting	A070103	Intermittent Capture	25/01/2023	25/01/2023
							F1905	Device Revision or Replacement	A070103	Intermittent Capture	25/01/2023	25/01/2023
					CT321	Long-term surgical invasive	E011903	Syncope/Fainting	A070103	Intermittent Capture	25/01/2023	25/01/2023
							F1905	Device Revision or Replacement	A070103	Intermittent Capture	25/01/2023	25/01/2023
					CT449	Generators and associated devices	E011903	Syncope/Fainting	A070103	Intermittent Capture	25/01/2023	25/01/2023
							F1905	Device Revision or Replacement	A070103	Intermittent Capture	25/01/2023	25/01/2023
					CT981	Single use	E011903	Syncope/Fainting	A070103	Intermittent Capture	25/01/2023	25/01/2023
							F1905	Device Revision or Replacement	A070103	Intermittent Capture	25/01/2023	25/01/2023
					CT983	Surgical invasive	E011903	Syncope/Fainting	A070103	Intermittent Capture	25/01/2023	25/01/2023
							F1905	Device Revision or Replacement	A070103	Intermittent Capture	25/01/2023	25/01/2023
13440	2023	All other reportable incidents	35852	Single-chamber implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	03/10/2023	03/10/2023
									A0908	Incorrect, Inadequate or Imprecise Result or Readings	03/10/2023	03/10/2023
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	03/10/2023	03/10/2023
									A0908	Incorrect, Inadequate or Imprecise Result or Readings	03/10/2023	03/10/2023
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	03/10/2023	03/10/2023
									A0908	Incorrect, Inadequate or Imprecise Result or Readings	03/10/2023	03/10/2023
13441	2023	All other reportable incidents	35852	Single-chamber implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	31/08/2023	31/08/2023
									A0908	Incorrect, Inadequate or Imprecise Result or Readings	31/08/2023	31/08/2023
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	31/08/2023	31/08/2023
									A0908	Incorrect, Inadequate or Imprecise Result or Readings	31/08/2023	31/08/2023
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	31/08/2023	31/08/2023
									A0908	Incorrect, Inadequate or Imprecise Result or Readings	31/08/2023	31/08/2023
13442	2023	All other reportable incidents	47265	Dual-chamber implantable pacemaker, rate-responsive	CT141	Electrodes/Leads and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A0203	Defective Device		
							F26	No Health Consequences or Impact	A0203	Defective Device		
					CT145	Cardiovascular devices	E2403	No Clinical Signs, Symptoms or Conditions	A0203	Defective Device		
							F26	No Health Consequences or Impact	A0203	Defective Device		
					CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A0203	Defective Device		
							F26	No Health Consequences or Impact	A0203	Defective Device		
					CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A0203	Defective Device		
							F26	No Health Consequences or Impact	A0203	Defective Device		
					CT233	Surgical	E2403	No Clinical Signs, Symptoms or Conditions	A0203	Defective Device		
							F26	No Health Consequences or Impact	A0203	Defective Device		
					CT2406	Implantable	E2403	No Clinical Signs, Symptoms or Conditions	A0203	Defective Device		
							F26	No Health Consequences or Impact	A0203	Defective Device		
					CT2407	Active implantable	E2403	No Clinical Signs, Symptoms or Conditions	A0203	Defective Device		
							F26	No Health Consequences or Impact	A0203	Defective Device		
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A0203	Defective Device		
							F26	No Health Consequences or Impact	A0203	Defective Device		
					CT2727	Active implant devices/ancillaries	E2403	No Clinical Signs, Symptoms or Conditions	A0203	Defective Device		
							F26	No Health Consequences or Impact	A0203	Defective Device		

					CT2750	Implantable cardiac electrical stimulators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A0203	Defective Device		
							F26	No Health Consequences or Impact	A0203	Defective Device		
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A0203	Defective Device		
							F26	No Health Consequences or Impact	A0203	Defective Device		
					CT2985	Active	E2403	No Clinical Signs, Symptoms or Conditions	A0203	Defective Device		
							F26	No Health Consequences or Impact	A0203	Defective Device		
					CT3001	Implantable metal	E2403	No Clinical Signs, Symptoms or Conditions	A0203	Defective Device		
							F26	No Health Consequences or Impact	A0203	Defective Device		
					CT3002	Implantable synthetic polymer	E2403	No Clinical Signs, Symptoms or Conditions	A0203	Defective Device		
							F26	No Health Consequences or Impact	A0203	Defective Device		
					CT3006	Implantable inorganic material	E2403	No Clinical Signs, Symptoms or Conditions	A0203	Defective Device		
							F26	No Health Consequences or Impact	A0203	Defective Device		
					CT314	Electrically powered	E2403	No Clinical Signs, Symptoms or Conditions	A0203	Defective Device		
							F26	No Health Consequences or Impact	A0203	Defective Device		
					CT321	Long-term surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A0203	Defective Device		
							F26	No Health Consequences or Impact	A0203	Defective Device		
					CT449	Generators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A0203	Defective Device		
							F26	No Health Consequences or Impact	A0203	Defective Device		
					CT981	Single use	E2403	No Clinical Signs, Symptoms or Conditions	A0203	Defective Device		
							F26	No Health Consequences or Impact	A0203	Defective Device		
					CT983	Surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A0203	Defective Device		
							F26	No Health Consequences or Impact	A0203	Defective Device		
13443	2023		47265	Dual-chamber implantable pacemaker, rate-responsive	CT141	Electrodes/Leads and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	16/06/2023	16/06/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	16/06/2023	16/06/2023
					CT145	Cardiovascular devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	16/06/2023	16/06/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	16/06/2023	16/06/2023
					CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	16/06/2023	16/06/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	16/06/2023	16/06/2023
					CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	16/06/2023	16/06/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	16/06/2023	16/06/2023
					CT233	Surgical	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	16/06/2023	16/06/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	16/06/2023	16/06/2023
					CT2406	Implantable	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	16/06/2023	16/06/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	16/06/2023	16/06/2023
					CT2407	Active implantable	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	16/06/2023	16/06/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	16/06/2023	16/06/2023
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	16/06/2023	16/06/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	16/06/2023	16/06/2023
					CT2727	Active implant devices/ancillaries	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	16/06/2023	16/06/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	16/06/2023	16/06/2023
					CT2750	Implantable cardiac electrical stimulators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	16/06/2023	16/06/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	16/06/2023	16/06/2023
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	16/06/2023	16/06/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	16/06/2023	16/06/2023
					CT2985	Active	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	16/06/2023	16/06/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	16/06/2023	16/06/2023
					CT3001	Implantable metal	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	16/06/2023	16/06/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	16/06/2023	16/06/2023
					CT3002	Implantable synthetic polymer	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	16/06/2023	16/06/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	16/06/2023	16/06/2023
					CT3006	Implantable inorganic material	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	16/06/2023	16/06/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	16/06/2023	16/06/2023
					CT314	Electrically powered	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	16/06/2023	16/06/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	16/06/2023	16/06/2023
					CT321	Long-term surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	16/06/2023	16/06/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	16/06/2023	16/06/2023
					CT449	Generators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	16/06/2023	16/06/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	16/06/2023	16/06/2023
					CT981	Single use	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	16/06/2023	16/06/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	16/06/2023	16/06/2023
					CT983	Surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	16/06/2023	16/06/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	16/06/2023	16/06/2023
13444	2023		47265	Dual-chamber implantable pacemaker, rate-responsive	CT141	Electrodes/Leads and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	15/06/2023	15/06/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	15/06/2023	15/06/2023
					CT145	Cardiovascular devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	15/06/2023	15/06/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	15/06/2023	15/06/2023
					CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	15/06/2023	15/06/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	15/06/2023	15/06/2023
					CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	15/06/2023	15/06/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	15/06/2023	15/06/2023
					CT233	Surgical	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	15/06/2023	15/06/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	15/06/2023	15/06/2023
					CT2406	Implantable	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	15/06/2023	15/06/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	15/06/2023	15/06/2023
					CT2407	Active implantable	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	15/06/2023	15/06/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	15/06/2023	15/06/2023
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	15/06/2023	15/06/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	15/06/2023	15/06/2023
					CT2727	Active implant devices/ancillaries	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	15/06/2023	15/06/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	15/06/2023	15/06/2023
					CT2750	Implantable cardiac electrical stimulators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	15/06/2023	15/06/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	15/06/2023	15/06/2023
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	15/06/2023	15/06/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	15/06/2023	15/06/2023
					CT2985	Active	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	15/06/2023	15/06/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	15/06/2023	15/06/2023
					CT3001	Implantable metal	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	15/06/2023	15/06/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	15/06/2023	15/06/2023

					CT3002	Implantable synthetic polymer	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	15/06/2023	15/06/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	15/06/2023	15/06/2023
					CT3006	Implantable inorganic material	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	15/06/2023	15/06/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	15/06/2023	15/06/2023
					CT314	Electrically powered	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	15/06/2023	15/06/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	15/06/2023	15/06/2023
					CT321	Long-term surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	15/06/2023	15/06/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	15/06/2023	15/06/2023
					CT449	Generators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	15/06/2023	15/06/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	15/06/2023	15/06/2023
					CT981	Single use	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	15/06/2023	15/06/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	15/06/2023	15/06/2023
					CT983	Surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	15/06/2023	15/06/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	15/06/2023	15/06/2023
13445	2023		47265	Dual-chamber implantable pacemaker, rate-responsive	CT141	Electrodes/Leads and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	16/06/2023	16/06/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	16/06/2023	16/06/2023
					CT145	Cardiovascular devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	16/06/2023	16/06/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	16/06/2023	16/06/2023
					CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	16/06/2023	16/06/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	16/06/2023	16/06/2023
					CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	16/06/2023	16/06/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	16/06/2023	16/06/2023
					CT233	Surgical	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	16/06/2023	16/06/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	16/06/2023	16/06/2023
					CT2406	Implantable	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	16/06/2023	16/06/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	16/06/2023	16/06/2023
					CT2407	Active implantable	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	16/06/2023	16/06/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	16/06/2023	16/06/2023
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	16/06/2023	16/06/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	16/06/2023	16/06/2023
					CT2727	Active implant devices/ancillaries	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	16/06/2023	16/06/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	16/06/2023	16/06/2023
					CT2750	Implantable cardiac electrical stimulators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	16/06/2023	16/06/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	16/06/2023	16/06/2023
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	16/06/2023	16/06/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	16/06/2023	16/06/2023
					CT2985	Active	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	16/06/2023	16/06/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	16/06/2023	16/06/2023
					CT3001	Implantable metal	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	16/06/2023	16/06/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	16/06/2023	16/06/2023
					CT3002	Implantable synthetic polymer	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	16/06/2023	16/06/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	16/06/2023	16/06/2023
					CT3006	Implantable inorganic material	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	16/06/2023	16/06/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	16/06/2023	16/06/2023
					CT314	Electrically powered	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	16/06/2023	16/06/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	16/06/2023	16/06/2023
					CT321	Long-term surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	16/06/2023	16/06/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	16/06/2023	16/06/2023
					CT449	Generators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	16/06/2023	16/06/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	16/06/2023	16/06/2023
					CT981	Single use	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	16/06/2023	16/06/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	16/06/2023	16/06/2023
					CT983	Surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	16/06/2023	16/06/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	16/06/2023	16/06/2023
13446	2023	All other reportable incidents	47265	Dual-chamber implantable pacemaker, rate-responsive	CT141	Electrodes/Leads and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	16/06/2023	16/06/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	16/06/2023	16/06/2023
					CT145	Cardiovascular devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	16/06/2023	16/06/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	16/06/2023	16/06/2023
					CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	16/06/2023	16/06/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	16/06/2023	16/06/2023
					CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	16/06/2023	16/06/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	16/06/2023	16/06/2023
					CT233	Surgical	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	16/06/2023	16/06/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	16/06/2023	16/06/2023
					CT2406	Implantable	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	16/06/2023	16/06/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	16/06/2023	16/06/2023
					CT2407	Active implantable	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	16/06/2023	16/06/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	16/06/2023	16/06/2023
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	16/06/2023	16/06/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	16/06/2023	16/06/2023
					CT2727	Active implant devices/ancillaries	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	16/06/2023	16/06/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	16/06/2023	16/06/2023
					CT2750	Implantable cardiac electrical stimulators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	16/06/2023	16/06/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	16/06/2023	16/06/2023
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	16/06/2023	16/06/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	16/06/2023	16/06/2023
					CT2985	Active	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	16/06/2023	16/06/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	16/06/2023	16/06/2023
					CT3001	Implantable metal	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	16/06/2023	16/06/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	16/06/2023	16/06/2023
					CT3002	Implantable synthetic polymer	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	16/06/2023	16/06/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	16/06/2023	16/06/2023
					CT3006	Implantable inorganic material	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	16/06/2023	16/06/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	16/06/2023	16/06/2023
					CT314	Electrically powered	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	16/06/2023	16/06/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	16/06/2023	16/06/2023
					CT321	Long-term surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	16/06/2023	16/06/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	16/06/2023	16/06/2023

					CT449	Generators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	16/06/2023	16/06/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	16/06/2023	16/06/2023
					CT981	Single use	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	16/06/2023	16/06/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	16/06/2023	16/06/2023
					CT983	Surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	16/06/2023	16/06/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	16/06/2023	16/06/2023
13447	2023	All other reportable incidents	47270	Cardiac resynchronization therapy implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	20/06/2023	20/06/2023
							F1901	Additional Surgery	A070504	Premature Discharge of Battery	20/06/2023	20/06/2023
13448	2023	All other reportable incidents	47265	Dual-chamber implantable pacemaker, rate-responsive	CT141	Electrodes/Leads and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT145	Cardiovascular devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT233	Surgical	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT2406	Implantable	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT2407	Active implantable	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT2727	Active implant devices/ancillaries	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT2750	Implantable cardiac electrical stimulators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT2985	Active	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT3001	Implantable metal	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT3002	Implantable synthetic polymer	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT3006	Implantable inorganic material	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT314	Electrically powered	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT321	Long-term surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT449	Generators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT981	Single use	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT983	Surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
13449	2023	All other reportable incidents	47265	Dual-chamber implantable pacemaker, rate-responsive	CT141	Electrodes/Leads and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	07/08/2023	07/08/2023
									A071102	Failure to Interrogate	07/08/2023	07/08/2023
							F1901	Additional Surgery	A070504	Premature Discharge of Battery	07/08/2023	07/08/2023
									A071102	Failure to Interrogate	07/08/2023	07/08/2023
					CT145	Cardiovascular devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	07/08/2023	07/08/2023
									A071102	Failure to Interrogate	07/08/2023	07/08/2023
							F1901	Additional Surgery	A070504	Premature Discharge of Battery	07/08/2023	07/08/2023
									A071102	Failure to Interrogate	07/08/2023	07/08/2023
					CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	07/08/2023	07/08/2023
									A071102	Failure to Interrogate	07/08/2023	07/08/2023
							F1901	Additional Surgery	A070504	Premature Discharge of Battery	07/08/2023	07/08/2023
									A071102	Failure to Interrogate	07/08/2023	07/08/2023
					CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	07/08/2023	07/08/2023
									A071102	Failure to Interrogate	07/08/2023	07/08/2023
							F1901	Additional Surgery	A070504	Premature Discharge of Battery	07/08/2023	07/08/2023
									A071102	Failure to Interrogate	07/08/2023	07/08/2023
					CT233	Surgical	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	07/08/2023	07/08/2023
									A071102	Failure to Interrogate	07/08/2023	07/08/2023
							F1901	Additional Surgery	A070504	Premature Discharge of Battery	07/08/2023	07/08/2023
									A071102	Failure to Interrogate	07/08/2023	07/08/2023
					CT2406	Implantable	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	07/08/2023	07/08/2023
									A071102	Failure to Interrogate	07/08/2023	07/08/2023
							F1901	Additional Surgery	A070504	Premature Discharge of Battery	07/08/2023	07/08/2023
									A071102	Failure to Interrogate	07/08/2023	07/08/2023
					CT2407	Active implantable	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	07/08/2023	07/08/2023
									A071102	Failure to Interrogate	07/08/2023	07/08/2023
							F1901	Additional Surgery	A070504	Premature Discharge of Battery	07/08/2023	07/08/2023
									A071102	Failure to Interrogate	07/08/2023	07/08/2023
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	07/08/2023	07/08/2023
									A071102	Failure to Interrogate	07/08/2023	07/08/2023
							F1901	Additional Surgery	A070504	Premature Discharge of Battery	07/08/2023	07/08/2023
									A071102	Failure to Interrogate	07/08/2023	07/08/2023
					CT2727	Active implant devices/ancillaries	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	07/08/2023	07/08/2023
									A071102	Failure to Interrogate	07/08/2023	07/08/2023
							F1901	Additional Surgery	A070504	Premature Discharge of Battery	07/08/2023	07/08/2023
									A071102	Failure to Interrogate	07/08/2023	07/08/2023
					CT2750	Implantable cardiac electrical stimulators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	07/08/2023	07/08/2023
									A071102	Failure to Interrogate	07/08/2023	07/08/2023
							F1901	Additional Surgery	A070504	Premature Discharge of Battery	07/08/2023	07/08/2023
									A071102	Failure to Interrogate	07/08/2023	07/08/2023

					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	07/08/2023	07/08/2023
									A071102	Failure to Interrogate	07/08/2023	07/08/2023
							F1901	Additional Surgery	A070504	Premature Discharge of Battery	07/08/2023	07/08/2023
									A071102	Failure to Interrogate	07/08/2023	07/08/2023
					CT2985	Active	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	07/08/2023	07/08/2023
									A071102	Failure to Interrogate	07/08/2023	07/08/2023
							F1901	Additional Surgery	A070504	Premature Discharge of Battery	07/08/2023	07/08/2023
									A071102	Failure to Interrogate	07/08/2023	07/08/2023
					CT3001	Implantable metal	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	07/08/2023	07/08/2023
									A071102	Failure to Interrogate	07/08/2023	07/08/2023
							F1901	Additional Surgery	A070504	Premature Discharge of Battery	07/08/2023	07/08/2023
									A071102	Failure to Interrogate	07/08/2023	07/08/2023
					CT3002	Implantable synthetic polymer	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	07/08/2023	07/08/2023
									A071102	Failure to Interrogate	07/08/2023	07/08/2023
							F1901	Additional Surgery	A070504	Premature Discharge of Battery	07/08/2023	07/08/2023
									A071102	Failure to Interrogate	07/08/2023	07/08/2023
					CT3006	Implantable inorganic material	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	07/08/2023	07/08/2023
									A071102	Failure to Interrogate	07/08/2023	07/08/2023
							F1901	Additional Surgery	A070504	Premature Discharge of Battery	07/08/2023	07/08/2023
									A071102	Failure to Interrogate	07/08/2023	07/08/2023
					CT314	Electrically powered	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	07/08/2023	07/08/2023
									A071102	Failure to Interrogate	07/08/2023	07/08/2023
							F1901	Additional Surgery	A070504	Premature Discharge of Battery	07/08/2023	07/08/2023
									A071102	Failure to Interrogate	07/08/2023	07/08/2023
					CT321	Long-term surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	07/08/2023	07/08/2023
									A071102	Failure to Interrogate	07/08/2023	07/08/2023
							F1901	Additional Surgery	A070504	Premature Discharge of Battery	07/08/2023	07/08/2023
									A071102	Failure to Interrogate	07/08/2023	07/08/2023
					CT449	Generators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	07/08/2023	07/08/2023
									A071102	Failure to Interrogate	07/08/2023	07/08/2023
							F1901	Additional Surgery	A070504	Premature Discharge of Battery	07/08/2023	07/08/2023
									A071102	Failure to Interrogate	07/08/2023	07/08/2023
					CT981	Single use	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	07/08/2023	07/08/2023
									A071102	Failure to Interrogate	07/08/2023	07/08/2023
							F1901	Additional Surgery	A070504	Premature Discharge of Battery	07/08/2023	07/08/2023
									A071102	Failure to Interrogate	07/08/2023	07/08/2023
					CT983	Surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	07/08/2023	07/08/2023
									A071102	Failure to Interrogate	07/08/2023	07/08/2023
							F1901	Additional Surgery	A070504	Premature Discharge of Battery	07/08/2023	07/08/2023
									A071102	Failure to Interrogate	07/08/2023	07/08/2023
13450	2023	All other reportable incidents	47205	Cardiac pulse generator programmer	CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A0203	Defective Device		
							F24	Insufficient Health Impact Information	A0203	Defective Device		
13451	2023		47265	Dual-chamber implantable pacemaker, rate-responsive	CT141	Electrodes/Leads and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	08/06/2023	08/06/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	08/06/2023	08/06/2023
					CT145	Cardiovascular devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	08/06/2023	08/06/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	08/06/2023	08/06/2023
					CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	08/06/2023	08/06/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	08/06/2023	08/06/2023
					CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	08/06/2023	08/06/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	08/06/2023	08/06/2023
					CT233	Surgical	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	08/06/2023	08/06/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	08/06/2023	08/06/2023
					CT2406	Implantable	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	08/06/2023	08/06/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	08/06/2023	08/06/2023
					CT2407	Active implantable	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	08/06/2023	08/06/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	08/06/2023	08/06/2023
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	08/06/2023	08/06/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	08/06/2023	08/06/2023
					CT2727	Active implant devices/ancillaries	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	08/06/2023	08/06/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	08/06/2023	08/06/2023
					CT2750	Implantable cardiac electrical stimulators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	08/06/2023	08/06/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	08/06/2023	08/06/2023
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	08/06/2023	08/06/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	08/06/2023	08/06/2023
					CT2985	Active	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	08/06/2023	08/06/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	08/06/2023	08/06/2023
					CT3001	Implantable metal	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	08/06/2023	08/06/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	08/06/2023	08/06/2023
					CT3002	Implantable synthetic polymer	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	08/06/2023	08/06/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	08/06/2023	08/06/2023
					CT3006	Implantable inorganic material	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	08/06/2023	08/06/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	08/06/2023	08/06/2023
					CT314	Electrically powered	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	08/06/2023	08/06/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	08/06/2023	08/06/2023
					CT321	Long-term surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	08/06/2023	08/06/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	08/06/2023	08/06/2023
					CT449	Generators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	08/06/2023	08/06/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	08/06/2023	08/06/2023
					CT981	Single use	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	08/06/2023	08/06/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	08/06/2023	08/06/2023
					CT983	Surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	08/06/2023	08/06/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	08/06/2023	08/06/2023
13452	2023	All other reportable incidents	17882	Physiologic-monitoring defibrillation system	CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A07	Electrical /Electronic Property Problem		
							F27	Problem Identified During Non-Clinical Procedure	A07	Electrical /Electronic Property Problem		
13453	2023	All other reportable incidents	47265	Dual-chamber implantable pacemaker, rate-responsive	CT141	Electrodes/Leads and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	13/03/2023	13/03/2023
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	13/03/2023	13/03/2023
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	13/03/2023	13/03/2023
					CT145	Cardiovascular devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	13/03/2023	13/03/2023

							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	13/03/2023	13/03/2023
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	13/03/2023	13/03/2023
					CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	13/03/2023	13/03/2023
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	13/03/2023	13/03/2023
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	13/03/2023	13/03/2023
					CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	13/03/2023	13/03/2023
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	13/03/2023	13/03/2023
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	13/03/2023	13/03/2023
					CT233	Surgical	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	13/03/2023	13/03/2023
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	13/03/2023	13/03/2023
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	13/03/2023	13/03/2023
					CT2406	Implantable	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	13/03/2023	13/03/2023
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	13/03/2023	13/03/2023
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	13/03/2023	13/03/2023
					CT2407	Active implantable	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	13/03/2023	13/03/2023
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	13/03/2023	13/03/2023
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	13/03/2023	13/03/2023
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	13/03/2023	13/03/2023
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	13/03/2023	13/03/2023
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	13/03/2023	13/03/2023
					CT2727	Active implant devices/ancillaries	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	13/03/2023	13/03/2023
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	13/03/2023	13/03/2023
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	13/03/2023	13/03/2023
					CT2750	Implantable cardiac electrical stimulators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	13/03/2023	13/03/2023
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	13/03/2023	13/03/2023
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	13/03/2023	13/03/2023
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	13/03/2023	13/03/2023
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	13/03/2023	13/03/2023
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	13/03/2023	13/03/2023
					CT2985	Active	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	13/03/2023	13/03/2023
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	13/03/2023	13/03/2023
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	13/03/2023	13/03/2023
					CT3001	Implantable metal	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	13/03/2023	13/03/2023
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	13/03/2023	13/03/2023
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	13/03/2023	13/03/2023
					CT3002	Implantable synthetic polymer	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	13/03/2023	13/03/2023
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	13/03/2023	13/03/2023
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	13/03/2023	13/03/2023
					CT3006	Implantable inorganic material	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	13/03/2023	13/03/2023
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	13/03/2023	13/03/2023
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	13/03/2023	13/03/2023
					CT314	Electrically powered	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	13/03/2023	13/03/2023
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	13/03/2023	13/03/2023
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	13/03/2023	13/03/2023
					CT321	Long-term surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	13/03/2023	13/03/2023
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	13/03/2023	13/03/2023
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	13/03/2023	13/03/2023
					CT449	Generators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	13/03/2023	13/03/2023
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	13/03/2023	13/03/2023
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	13/03/2023	13/03/2023
					CT981	Single use	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	13/03/2023	13/03/2023
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	13/03/2023	13/03/2023
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	13/03/2023	13/03/2023
					CT983	Surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	13/03/2023	13/03/2023
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	13/03/2023	13/03/2023
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	13/03/2023	13/03/2023
13454	2023	All other reportable incidents	47265	Dual-chamber implantable pacemaker, rate-responsive	CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
							F08	Hospitalization or Prolonged Hospitalization	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
							F1903	Device Explantation	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
							F08	Hospitalization or Prolonged Hospitalization	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
							F1903	Device Explantation	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
							F08	Hospitalization or Prolonged Hospitalization	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
							F1903	Device Explantation	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
13455	2023	All other reportable incidents	17882	Physiologic-monitoring defibrillation system	CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070803	Failure to Power Up		
							F27	Problem Identified During Non-Clinical Procedure	A070803	Failure to Power Up		
13456	2023	All other reportable incidents	17882	Physiologic-monitoring defibrillation system	CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A150105	Key or Button Unresponsive/not Working		
							F27	Problem Identified During Non-Clinical Procedure	A150105	Key or Button Unresponsive/not Working		
13457	2023	All other reportable incidents	47270	Cardiac resynchronization therapy implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	23/01/2023	23/01/2023
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	23/01/2023	23/01/2023
							F12	Serious Injury/ Illness/ Impairment	A070504	Premature Discharge of Battery	23/01/2023	23/01/2023
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	23/01/2023	23/01/2023
13458	2023	All other reportable incidents	37265	Dual-chamber implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	23/06/2023	23/06/2023
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	23/06/2023	23/06/2023
							F12	Serious Injury/ Illness/ Impairment	A070504	Premature Discharge of Battery	23/06/2023	23/06/2023
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	23/06/2023	23/06/2023
13459	2023	All other reportable incidents	47205	Cardiac pulse generator programmer	CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A0203	Defective Device		
							F24	Insufficient Health Impact Information	A0203	Defective Device		

13460	2023	All other reportable incidents	17882	Physiologic-monitoring defibrillation system	CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A0801	Failure to Calibrate		
							F27	Problem Identified During Non-Clinical Procedure	A0801	Failure to Calibrate		
13461	2023	All other reportable incidents	35223	Endocardial/interventricular septal pacing lead	CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070101	Failure to Capture		
									A070102	High Capture Threshold		
									A070104	Unstable Capture Threshold		
							F08	Hospitalization or Prolonged Hospitalization	A070101	Failure to Capture		
									A070102	High Capture Threshold		
									A070104	Unstable Capture Threshold		
							F12	Serious Injury/ Illness/ Impairment	A070101	Failure to Capture		
									A070102	High Capture Threshold		
									A070104	Unstable Capture Threshold		
							F1905	Device Revision or Replacement	A070101	Failure to Capture		
									A070102	High Capture Threshold		
									A070104	Unstable Capture Threshold		
							F23	Unexpected Medical Intervention	A070101	Failure to Capture		
									A070102	High Capture Threshold		
									A070104	Unstable Capture Threshold		
13462	2023		47265	Dual-chamber implantable pacemaker, rate-responsive	CT141	Electrodes/Leads and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	21/05/2023	21/05/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	21/05/2023	21/05/2023
					CT145	Cardiovascular devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	21/05/2023	21/05/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	21/05/2023	21/05/2023
					CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	21/05/2023	21/05/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	21/05/2023	21/05/2023
					CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	21/05/2023	21/05/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	21/05/2023	21/05/2023
					CT233	Surgical	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	21/05/2023	21/05/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	21/05/2023	21/05/2023
					CT2406	Implantable	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	21/05/2023	21/05/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	21/05/2023	21/05/2023
					CT2407	Active implantable	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	21/05/2023	21/05/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	21/05/2023	21/05/2023
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	21/05/2023	21/05/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	21/05/2023	21/05/2023
					CT2727	Active implant devices/ancillaries	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	21/05/2023	21/05/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	21/05/2023	21/05/2023
					CT2750	Implantable cardiac electrical stimulators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	21/05/2023	21/05/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	21/05/2023	21/05/2023
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	21/05/2023	21/05/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	21/05/2023	21/05/2023
					CT2985	Active	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	21/05/2023	21/05/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	21/05/2023	21/05/2023
					CT3001	Implantable metal	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	21/05/2023	21/05/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	21/05/2023	21/05/2023
					CT3002	Implantable synthetic polymer	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	21/05/2023	21/05/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	21/05/2023	21/05/2023
					CT3006	Implantable inorganic material	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	21/05/2023	21/05/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	21/05/2023	21/05/2023
					CT314	Electrically powered	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	21/05/2023	21/05/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	21/05/2023	21/05/2023
					CT321	Long-term surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	21/05/2023	21/05/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	21/05/2023	21/05/2023
					CT449	Generators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	21/05/2023	21/05/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	21/05/2023	21/05/2023
					CT981	Single use	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	21/05/2023	21/05/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	21/05/2023	21/05/2023
					CT983	Surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	21/05/2023	21/05/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	21/05/2023	21/05/2023
13463	2023		47265	Dual-chamber implantable pacemaker, rate-responsive	CT141	Electrodes/Leads and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	21/05/2023	21/05/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	21/05/2023	21/05/2023
					CT145	Cardiovascular devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	21/05/2023	21/05/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	21/05/2023	21/05/2023
					CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	21/05/2023	21/05/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	21/05/2023	21/05/2023
					CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	21/05/2023	21/05/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	21/05/2023	21/05/2023
					CT233	Surgical	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	21/05/2023	21/05/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	21/05/2023	21/05/2023
					CT2406	Implantable	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	21/05/2023	21/05/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	21/05/2023	21/05/2023
					CT2407	Active implantable	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	21/05/2023	21/05/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	21/05/2023	21/05/2023
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	21/05/2023	21/05/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	21/05/2023	21/05/2023
					CT2727	Active implant devices/ancillaries	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	21/05/2023	21/05/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	21/05/2023	21/05/2023
					CT2750	Implantable cardiac electrical stimulators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	21/05/2023	21/05/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	21/05/2023	21/05/2023
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	21/05/2023	21/05/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	21/05/2023	21/05/2023
					CT2985	Active	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	21/05/2023	21/05/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	21/05/2023	21/05/2023
					CT3001	Implantable metal	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	21/05/2023	21/05/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	21/05/2023	21/05/2023
					CT3002	Implantable synthetic polymer	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	21/05/2023	21/05/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	21/05/2023	21/05/2023
					CT3006	Implantable inorganic material	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	21/05/2023	21/05/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	21/05/2023	21/05/2023
					CT314	Electrically powered	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	21/05/2023	21/05/2023

							F1901	Additional Surgery	A24	No Apparent Device Problem	21/05/2023	21/05/2023
					CT321	Long-term surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	21/05/2023	21/05/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	21/05/2023	21/05/2023
					CT449	Generators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	21/05/2023	21/05/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	21/05/2023	21/05/2023
					CT981	Single use	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	21/05/2023	21/05/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	21/05/2023	21/05/2023
					CT983	Surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	21/05/2023	21/05/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	21/05/2023	21/05/2023
13464	2023	All other reportable incidents	47265	Dual-chamber implantable pacemaker, rate-responsive	CT141	Electrodes/Leads and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070909	Oversensing		
							F26	No Health Consequences or Impact	A070909	Oversensing		
					CT145	Cardiovascular devices	E2403	No Clinical Signs, Symptoms or Conditions	A070909	Oversensing		
							F26	No Health Consequences or Impact	A070909	Oversensing		
					CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070909	Oversensing		
							F26	No Health Consequences or Impact	A070909	Oversensing		
					CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A070909	Oversensing		
							F26	No Health Consequences or Impact	A070909	Oversensing		
					CT233	Surgical	E2403	No Clinical Signs, Symptoms or Conditions	A070909	Oversensing		
							F26	No Health Consequences or Impact	A070909	Oversensing		
					CT2406	Implantable	E2403	No Clinical Signs, Symptoms or Conditions	A070909	Oversensing		
							F26	No Health Consequences or Impact	A070909	Oversensing		
					CT2407	Active implantable	E2403	No Clinical Signs, Symptoms or Conditions	A070909	Oversensing		
							F26	No Health Consequences or Impact	A070909	Oversensing		
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A070909	Oversensing		
							F26	No Health Consequences or Impact	A070909	Oversensing		
					CT2727	Active implant devices/ancillaries	E2403	No Clinical Signs, Symptoms or Conditions	A070909	Oversensing		
							F26	No Health Consequences or Impact	A070909	Oversensing		
					CT2750	Implantable cardiac electrical stimulators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070909	Oversensing		
							F26	No Health Consequences or Impact	A070909	Oversensing		
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A070909	Oversensing		
							F26	No Health Consequences or Impact	A070909	Oversensing		
					CT2985	Active	E2403	No Clinical Signs, Symptoms or Conditions	A070909	Oversensing		
							F26	No Health Consequences or Impact	A070909	Oversensing		
					CT3001	Implantable metal	E2403	No Clinical Signs, Symptoms or Conditions	A070909	Oversensing		
							F26	No Health Consequences or Impact	A070909	Oversensing		
					CT3002	Implantable synthetic polymer	E2403	No Clinical Signs, Symptoms or Conditions	A070909	Oversensing		
							F26	No Health Consequences or Impact	A070909	Oversensing		
					CT3006	Implantable inorganic material	E2403	No Clinical Signs, Symptoms or Conditions	A070909	Oversensing		
							F26	No Health Consequences or Impact	A070909	Oversensing		
					CT314	Electrically powered	E2403	No Clinical Signs, Symptoms or Conditions	A070909	Oversensing		
							F26	No Health Consequences or Impact	A070909	Oversensing		
					CT321	Long-term surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A070909	Oversensing		
							F26	No Health Consequences or Impact	A070909	Oversensing		
					CT449	Generators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070909	Oversensing		
							F26	No Health Consequences or Impact	A070909	Oversensing		
					CT981	Single use	E2403	No Clinical Signs, Symptoms or Conditions	A070909	Oversensing		
							F26	No Health Consequences or Impact	A070909	Oversensing		
					CT983	Surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A070909	Oversensing		
							F26	No Health Consequences or Impact	A070909	Oversensing		
13465	2023	All other reportable incidents	47265	Dual-chamber implantable pacemaker, rate-responsive	CT141	Electrodes/Leads and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem		
									A072201	High Impedance		
									A1301	Failure to Read Input Signal		
							F26	No Health Consequences or Impact	A0705	Battery Problem		
									A072201	High Impedance		
									A1301	Failure to Read Input Signal		
					CT145	Cardiovascular devices	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem		
									A072201	High Impedance		
									A1301	Failure to Read Input Signal		
							F26	No Health Consequences or Impact	A0705	Battery Problem		
									A072201	High Impedance		
									A1301	Failure to Read Input Signal		
					CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem		
									A072201	High Impedance		
									A1301	Failure to Read Input Signal		
							F26	No Health Consequences or Impact	A0705	Battery Problem		
									A072201	High Impedance		
									A1301	Failure to Read Input Signal		
					CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem		
									A072201	High Impedance		
									A1301	Failure to Read Input Signal		
							F26	No Health Consequences or Impact	A0705	Battery Problem		
									A072201	High Impedance		
									A1301	Failure to Read Input Signal		
					CT233	Surgical	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem		
									A072201	High Impedance		
									A1301	Failure to Read Input Signal		
							F26	No Health Consequences or Impact	A0705	Battery Problem		
									A072201	High Impedance		
									A1301	Failure to Read Input Signal		
					CT2406	Implantable	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem		
									A072201	High Impedance		
									A1301	Failure to Read Input Signal		
							F26	No Health Consequences or Impact	A0705	Battery Problem		
									A072201	High Impedance		
									A1301	Failure to Read Input Signal		
					CT2407	Active implantable	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem		
									A072201	High Impedance		
									A1301	Failure to Read Input Signal		

							F26	No Health Consequences or Impact	A0705	Battery Problem		
									A072201	High Impedance		
									A1301	Failure to Read Input Signal		
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem		
									A072201	High Impedance		
									A1301	Failure to Read Input Signal		
							F26	No Health Consequences or Impact	A0705	Battery Problem		
									A072201	High Impedance		
									A1301	Failure to Read Input Signal		
					CT2727	Active implant devices/ancillaries	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem		
									A072201	High Impedance		
									A1301	Failure to Read Input Signal		
							F26	No Health Consequences or Impact	A0705	Battery Problem		
									A072201	High Impedance		
									A1301	Failure to Read Input Signal		
					CT2750	Implantable cardiac electrical stimulators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem		
									A072201	High Impedance		
									A1301	Failure to Read Input Signal		
							F26	No Health Consequences or Impact	A0705	Battery Problem		
									A072201	High Impedance		
									A1301	Failure to Read Input Signal		
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem		
									A072201	High Impedance		
									A1301	Failure to Read Input Signal		
							F26	No Health Consequences or Impact	A0705	Battery Problem		
									A072201	High Impedance		
									A1301	Failure to Read Input Signal		
					CT2985	Active	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem		
									A072201	High Impedance		
									A1301	Failure to Read Input Signal		
							F26	No Health Consequences or Impact	A0705	Battery Problem		
									A072201	High Impedance		
									A1301	Failure to Read Input Signal		
					CT3001	Implantable metal	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem		
									A072201	High Impedance		
									A1301	Failure to Read Input Signal		
							F26	No Health Consequences or Impact	A0705	Battery Problem		
									A072201	High Impedance		
									A1301	Failure to Read Input Signal		
					CT3002	Implantable synthetic polymer	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem		
									A072201	High Impedance		
									A1301	Failure to Read Input Signal		
							F26	No Health Consequences or Impact	A0705	Battery Problem		
									A072201	High Impedance		
									A1301	Failure to Read Input Signal		
					CT3006	Implantable inorganic material	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem		
									A072201	High Impedance		
									A1301	Failure to Read Input Signal		
							F26	No Health Consequences or Impact	A0705	Battery Problem		
									A072201	High Impedance		
									A1301	Failure to Read Input Signal		
					CT314	Electrically powered	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem		
									A072201	High Impedance		
									A1301	Failure to Read Input Signal		
							F26	No Health Consequences or Impact	A0705	Battery Problem		
									A072201	High Impedance		
									A1301	Failure to Read Input Signal		
					CT321	Long-term surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem		
									A072201	High Impedance		
									A1301	Failure to Read Input Signal		
							F26	No Health Consequences or Impact	A0705	Battery Problem		
									A072201	High Impedance		
									A1301	Failure to Read Input Signal		
					CT449	Generators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem		
									A072201	High Impedance		
									A1301	Failure to Read Input Signal		
							F26	No Health Consequences or Impact	A0705	Battery Problem		
									A072201	High Impedance		
									A1301	Failure to Read Input Signal		
					CT981	Single use	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem		
									A072201	High Impedance		
									A1301	Failure to Read Input Signal		
							F26	No Health Consequences or Impact	A0705	Battery Problem		
									A072201	High Impedance		
									A1301	Failure to Read Input Signal		
					CT983	Surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem		
									A072201	High Impedance		
									A1301	Failure to Read Input Signal		
							F26	No Health Consequences or Impact	A0705	Battery Problem		
									A072201	High Impedance		
									A1301	Failure to Read Input Signal		
13466	2023		47265	Dual-chamber implantable pacemaker, rate-responsive	CT141	Electrodes/Leads and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	21/05/2023	21/05/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	21/05/2023	21/05/2023
					CT145	Cardiovascular devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	21/05/2023	21/05/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	21/05/2023	21/05/2023
					CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	21/05/2023	21/05/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	21/05/2023	21/05/2023
					CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	21/05/2023	21/05/2023

							F1901	Additional Surgery	A24	No Apparent Device Problem	21/05/2023	21/05/2023
					CT233	Surgical	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	21/05/2023	21/05/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	21/05/2023	21/05/2023
					CT2406	Implantable	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	21/05/2023	21/05/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	21/05/2023	21/05/2023
					CT2407	Active implantable	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	21/05/2023	21/05/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	21/05/2023	21/05/2023
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	21/05/2023	21/05/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	21/05/2023	21/05/2023
					CT2727	Active implant devices/ancillaries	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	21/05/2023	21/05/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	21/05/2023	21/05/2023
					CT2750	Implantable cardiac electrical stimulators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	21/05/2023	21/05/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	21/05/2023	21/05/2023
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	21/05/2023	21/05/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	21/05/2023	21/05/2023
					CT2985	Active	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	21/05/2023	21/05/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	21/05/2023	21/05/2023
					CT3001	Implantable metal	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	21/05/2023	21/05/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	21/05/2023	21/05/2023
					CT3002	Implantable synthetic polymer	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	21/05/2023	21/05/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	21/05/2023	21/05/2023
					CT3006	Implantable inorganic material	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	21/05/2023	21/05/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	21/05/2023	21/05/2023
					CT314	Electrically powered	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	21/05/2023	21/05/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	21/05/2023	21/05/2023
					CT321	Long-term surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	21/05/2023	21/05/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	21/05/2023	21/05/2023
					CT449	Generators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	21/05/2023	21/05/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	21/05/2023	21/05/2023
					CT981	Single use	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	21/05/2023	21/05/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	21/05/2023	21/05/2023
					CT983	Surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	21/05/2023	21/05/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	21/05/2023	21/05/2023
13467	2023		47265	Dual-chamber implantable pacemaker, rate-responsive	CT141	Electrodes/Leads and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	21/05/2023	21/05/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	21/05/2023	21/05/2023
					CT145	Cardiovascular devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	21/05/2023	21/05/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	21/05/2023	21/05/2023
					CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	21/05/2023	21/05/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	21/05/2023	21/05/2023
					CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	21/05/2023	21/05/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	21/05/2023	21/05/2023
					CT233	Surgical	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	21/05/2023	21/05/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	21/05/2023	21/05/2023
					CT2406	Implantable	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	21/05/2023	21/05/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	21/05/2023	21/05/2023
					CT2407	Active implantable	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	21/05/2023	21/05/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	21/05/2023	21/05/2023
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	21/05/2023	21/05/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	21/05/2023	21/05/2023
					CT2727	Active implant devices/ancillaries	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	21/05/2023	21/05/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	21/05/2023	21/05/2023
					CT2750	Implantable cardiac electrical stimulators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	21/05/2023	21/05/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	21/05/2023	21/05/2023
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	21/05/2023	21/05/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	21/05/2023	21/05/2023
					CT2985	Active	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	21/05/2023	21/05/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	21/05/2023	21/05/2023
					CT3001	Implantable metal	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	21/05/2023	21/05/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	21/05/2023	21/05/2023
					CT3002	Implantable synthetic polymer	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	21/05/2023	21/05/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	21/05/2023	21/05/2023
					CT3006	Implantable inorganic material	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	21/05/2023	21/05/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	21/05/2023	21/05/2023
					CT314	Electrically powered	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	21/05/2023	21/05/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	21/05/2023	21/05/2023
					CT321	Long-term surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	21/05/2023	21/05/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	21/05/2023	21/05/2023
					CT449	Generators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	21/05/2023	21/05/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	21/05/2023	21/05/2023
					CT981	Single use	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	21/05/2023	21/05/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	21/05/2023	21/05/2023
					CT983	Surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	21/05/2023	21/05/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	21/05/2023	21/05/2023
13468	2023		47265	Dual-chamber implantable pacemaker, rate-responsive	CT141	Electrodes/Leads and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	21/05/2023	21/05/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	21/05/2023	21/05/2023
					CT145	Cardiovascular devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	21/05/2023	21/05/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	21/05/2023	21/05/2023
					CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	21/05/2023	21/05/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	21/05/2023	21/05/2023
					CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	21/05/2023	21/05/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	21/05/2023	21/05/2023
					CT233	Surgical	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	21/05/2023	21/05/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	21/05/2023	21/05/2023
					CT2406	Implantable	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	21/05/2023	21/05/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	21/05/2023	21/05/2023
					CT2407	Active implantable	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	21/05/2023	21/05/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	21/05/2023	21/05/2023
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	21/05/2023	21/05/2023

						F1901	Additional Surgery	A24	No Apparent Device Problem	21/05/2023	21/05/2023
				CT2727	Active implant devices/ancillaries	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	21/05/2023	21/05/2023
						F1901	Additional Surgery	A24	No Apparent Device Problem	21/05/2023	21/05/2023
				CT2750	Implantable cardiac electrical stimulators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	21/05/2023	21/05/2023
						F1901	Additional Surgery	A24	No Apparent Device Problem	21/05/2023	21/05/2023
				CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	21/05/2023	21/05/2023
						F1901	Additional Surgery	A24	No Apparent Device Problem	21/05/2023	21/05/2023
				CT2985	Active	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	21/05/2023	21/05/2023
						F1901	Additional Surgery	A24	No Apparent Device Problem	21/05/2023	21/05/2023
				CT3001	Implantable metal	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	21/05/2023	21/05/2023
						F1901	Additional Surgery	A24	No Apparent Device Problem	21/05/2023	21/05/2023
				CT3002	Implantable synthetic polymer	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	21/05/2023	21/05/2023
						F1901	Additional Surgery	A24	No Apparent Device Problem	21/05/2023	21/05/2023
				CT3006	Implantable inorganic material	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	21/05/2023	21/05/2023
						F1901	Additional Surgery	A24	No Apparent Device Problem	21/05/2023	21/05/2023
				CT314	Electrically powered	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	21/05/2023	21/05/2023
						F1901	Additional Surgery	A24	No Apparent Device Problem	21/05/2023	21/05/2023
				CT321	Long-term surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	21/05/2023	21/05/2023
						F1901	Additional Surgery	A24	No Apparent Device Problem	21/05/2023	21/05/2023
				CT449	Generators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	21/05/2023	21/05/2023
						F1901	Additional Surgery	A24	No Apparent Device Problem	21/05/2023	21/05/2023
				CT981	Single use	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	21/05/2023	21/05/2023
						F1901	Additional Surgery	A24	No Apparent Device Problem	21/05/2023	21/05/2023
				CT983	Surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	21/05/2023	21/05/2023
						F1901	Additional Surgery	A24	No Apparent Device Problem	21/05/2023	21/05/2023
13469	2023		47265	Dual-chamber implantable pacemaker, rate-responsive	CT141	Electrodes/Leads and associated devices	E011903	Syncope/Fainting	A070101	Failure to Capture	05/07/2023	05/07/2023
						F1901	Additional Surgery	A070101	Failure to Capture	05/07/2023	05/07/2023
				CT145	Cardiovascular devices	E011903	Syncope/Fainting	A070101	Failure to Capture	05/07/2023	05/07/2023
						F1901	Additional Surgery	A070101	Failure to Capture	05/07/2023	05/07/2023
				CT1684	Cardiac pacemakers and associated devices	E011903	Syncope/Fainting	A070101	Failure to Capture	05/07/2023	05/07/2023
						F1901	Additional Surgery	A070101	Failure to Capture	05/07/2023	05/07/2023
				CT1685	Cardiac pacemakers	E011903	Syncope/Fainting	A070101	Failure to Capture	05/07/2023	05/07/2023
						F1901	Additional Surgery	A070101	Failure to Capture	05/07/2023	05/07/2023
				CT233	Surgical	E011903	Syncope/Fainting	A070101	Failure to Capture	05/07/2023	05/07/2023
						F1901	Additional Surgery	A070101	Failure to Capture	05/07/2023	05/07/2023
				CT2406	Implantable	E011903	Syncope/Fainting	A070101	Failure to Capture	05/07/2023	05/07/2023
						F1901	Additional Surgery	A070101	Failure to Capture	05/07/2023	05/07/2023
				CT2407	Active implantable	E011903	Syncope/Fainting	A070101	Failure to Capture	05/07/2023	05/07/2023
						F1901	Additional Surgery	A070101	Failure to Capture	05/07/2023	05/07/2023
				CT257	Implantable cardiac pacemakers	E011903	Syncope/Fainting	A070101	Failure to Capture	05/07/2023	05/07/2023
						F1901	Additional Surgery	A070101	Failure to Capture	05/07/2023	05/07/2023
				CT2727	Active implant devices/ancillaries	E011903	Syncope/Fainting	A070101	Failure to Capture	05/07/2023	05/07/2023
						F1901	Additional Surgery	A070101	Failure to Capture	05/07/2023	05/07/2023
				CT2750	Implantable cardiac electrical stimulators and associated devices	E011903	Syncope/Fainting	A070101	Failure to Capture	05/07/2023	05/07/2023
						F1901	Additional Surgery	A070101	Failure to Capture	05/07/2023	05/07/2023
				CT2866	Implantable cardiac electrical stimulators	E011903	Syncope/Fainting	A070101	Failure to Capture	05/07/2023	05/07/2023
						F1901	Additional Surgery	A070101	Failure to Capture	05/07/2023	05/07/2023
				CT2985	Active	E011903	Syncope/Fainting	A070101	Failure to Capture	05/07/2023	05/07/2023
						F1901	Additional Surgery	A070101	Failure to Capture	05/07/2023	05/07/2023
				CT3001	Implantable metal	E011903	Syncope/Fainting	A070101	Failure to Capture	05/07/2023	05/07/2023
						F1901	Additional Surgery	A070101	Failure to Capture	05/07/2023	05/07/2023
				CT3002	Implantable synthetic polymer	E011903	Syncope/Fainting	A070101	Failure to Capture	05/07/2023	05/07/2023
						F1901	Additional Surgery	A070101	Failure to Capture	05/07/2023	05/07/2023
				CT3006	Implantable inorganic material	E011903	Syncope/Fainting	A070101	Failure to Capture	05/07/2023	05/07/2023
						F1901	Additional Surgery	A070101	Failure to Capture	05/07/2023	05/07/2023
				CT314	Electrically powered	E011903	Syncope/Fainting	A070101	Failure to Capture	05/07/2023	05/07/2023
						F1901	Additional Surgery	A070101	Failure to Capture	05/07/2023	05/07/2023
				CT321	Long-term surgical invasive	E011903	Syncope/Fainting	A070101	Failure to Capture	05/07/2023	05/07/2023
						F1901	Additional Surgery	A070101	Failure to Capture	05/07/2023	05/07/2023
				CT449	Generators and associated devices	E011903	Syncope/Fainting	A070101	Failure to Capture	05/07/2023	05/07/2023
						F1901	Additional Surgery	A070101	Failure to Capture	05/07/2023	05/07/2023
				CT981	Single use	E011903	Syncope/Fainting	A070101	Failure to Capture	05/07/2023	05/07/2023
						F1901	Additional Surgery	A070101	Failure to Capture	05/07/2023	05/07/2023
				CT983	Surgical invasive	E011903	Syncope/Fainting	A070101	Failure to Capture	05/07/2023	05/07/2023
						F1901	Additional Surgery	A070101	Failure to Capture	05/07/2023	05/07/2023
13470	2023	All other reportable incidents	47205	Cardiac pulse generator programmer	CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A0203	Defective Device		
									A25	No Apparent Adverse Event		
							F26	No Health Consequences or Impact	A0203	Defective Device		
									A25	No Apparent Adverse Event		
13471	2023		35223	Endocardial/interventricular septal pacing lead	CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070909	Oversensing		
									A0712	Pacing Problem		
							F26	No Health Consequences or Impact	A070909	Oversensing		
									A0712	Pacing Problem		
13472	2023		35223	Endocardial/interventricular septal pacing lead	CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070909	Oversensing		
							F26	No Health Consequences or Impact	A070909	Oversensing		
13473	2023		47265	Dual-chamber implantable pacemaker, rate-responsive	CT141	Electrodes/Leads and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	04/06/2023	04/06/2023
						F1901	Additional Surgery	A24	No Apparent Device Problem	04/06/2023	04/06/2023
				CT145	Cardiovascular devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	04/06/2023	04/06/2023
						F1901	Additional Surgery	A24	No Apparent Device Problem	04/06/2023	04/06/2023
				CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	04/06/2023	04/06/2023
						F1901	Additional Surgery	A24	No Apparent Device Problem	04/06/2023	04/06/2023
				CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	04/06/2023	04/06/2023
						F1901	Additional Surgery	A24	No Apparent Device Problem	04/06/2023	04/06/2023
				CT233	Surgical	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	04/06/2023	04/06/2023
						F1901	Additional Surgery	A24	No Apparent Device Problem	04/06/2023	04/06/2023
				CT2406	Implantable	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	04/06/2023	04/06/2023
						F1901	Additional Surgery	A24	No Apparent Device Problem	04/06/2023	04/06/2023
				CT2407	Active implantable	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	04/06/2023	04/06/2023

							F1901	Additional Surgery	A24	No Apparent Device Problem	04/06/2023	04/06/2023
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	04/06/2023	04/06/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	04/06/2023	04/06/2023
					CT2727	Active implant devices/ancillaries	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	04/06/2023	04/06/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	04/06/2023	04/06/2023
					CT2750	Implantable cardiac electrical stimulators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	04/06/2023	04/06/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	04/06/2023	04/06/2023
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	04/06/2023	04/06/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	04/06/2023	04/06/2023
					CT2985	Active	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	04/06/2023	04/06/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	04/06/2023	04/06/2023
					CT3001	Implantable metal	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	04/06/2023	04/06/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	04/06/2023	04/06/2023
					CT3002	Implantable synthetic polymer	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	04/06/2023	04/06/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	04/06/2023	04/06/2023
					CT3006	Implantable inorganic material	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	04/06/2023	04/06/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	04/06/2023	04/06/2023
					CT314	Electrically powered	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	04/06/2023	04/06/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	04/06/2023	04/06/2023
					CT321	Long-term surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	04/06/2023	04/06/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	04/06/2023	04/06/2023
					CT449	Generators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	04/06/2023	04/06/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	04/06/2023	04/06/2023
					CT981	Single use	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	04/06/2023	04/06/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	04/06/2023	04/06/2023
					CT983	Surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	04/06/2023	04/06/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	04/06/2023	04/06/2023
13474	2023	All other reportable incidents	47265	Dual-chamber implantable pacemaker, rate-responsive	CT141	Electrodes/Leads and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	09/06/2023	09/06/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	09/06/2023	09/06/2023
					CT145	Cardiovascular devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	09/06/2023	09/06/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	09/06/2023	09/06/2023
					CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	09/06/2023	09/06/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	09/06/2023	09/06/2023
					CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	09/06/2023	09/06/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	09/06/2023	09/06/2023
					CT233	Surgical	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	09/06/2023	09/06/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	09/06/2023	09/06/2023
					CT2406	Implantable	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	09/06/2023	09/06/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	09/06/2023	09/06/2023
					CT2407	Active implantable	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	09/06/2023	09/06/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	09/06/2023	09/06/2023
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	09/06/2023	09/06/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	09/06/2023	09/06/2023
					CT2727	Active implant devices/ancillaries	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	09/06/2023	09/06/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	09/06/2023	09/06/2023
					CT2750	Implantable cardiac electrical stimulators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	09/06/2023	09/06/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	09/06/2023	09/06/2023
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	09/06/2023	09/06/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	09/06/2023	09/06/2023
					CT2985	Active	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	09/06/2023	09/06/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	09/06/2023	09/06/2023
					CT3001	Implantable metal	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	09/06/2023	09/06/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	09/06/2023	09/06/2023
					CT3002	Implantable synthetic polymer	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	09/06/2023	09/06/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	09/06/2023	09/06/2023
					CT3006	Implantable inorganic material	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	09/06/2023	09/06/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	09/06/2023	09/06/2023
					CT314	Electrically powered	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	09/06/2023	09/06/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	09/06/2023	09/06/2023
					CT321	Long-term surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	09/06/2023	09/06/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	09/06/2023	09/06/2023
					CT449	Generators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	09/06/2023	09/06/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	09/06/2023	09/06/2023
					CT981	Single use	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	09/06/2023	09/06/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	09/06/2023	09/06/2023
					CT983	Surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	09/06/2023	09/06/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	09/06/2023	09/06/2023
13475	2023		47265	Dual-chamber implantable pacemaker, rate-responsive	CT141	Electrodes/Leads and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	26/05/2023	26/05/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	26/05/2023	26/05/2023
					CT145	Cardiovascular devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	26/05/2023	26/05/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	26/05/2023	26/05/2023
					CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	26/05/2023	26/05/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	26/05/2023	26/05/2023
					CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	26/05/2023	26/05/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	26/05/2023	26/05/2023
					CT233	Surgical	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	26/05/2023	26/05/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	26/05/2023	26/05/2023
					CT2406	Implantable	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	26/05/2023	26/05/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	26/05/2023	26/05/2023
					CT2407	Active implantable	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	26/05/2023	26/05/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	26/05/2023	26/05/2023
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	26/05/2023	26/05/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	26/05/2023	26/05/2023
					CT2727	Active implant devices/ancillaries	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	26/05/2023	26/05/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	26/05/2023	26/05/2023
					CT2750	Implantable cardiac electrical stimulators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	26/05/2023	26/05/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	26/05/2023	26/05/2023
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	26/05/2023	26/05/2023

							F1901	Additional Surgery	A24	No Apparent Device Problem	26/05/2023	26/05/2023
					CT2985	Active	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	26/05/2023	26/05/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	26/05/2023	26/05/2023
					CT3001	Implantable metal	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	26/05/2023	26/05/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	26/05/2023	26/05/2023
					CT3002	Implantable synthetic polymer	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	26/05/2023	26/05/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	26/05/2023	26/05/2023
					CT3006	Implantable inorganic material	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	26/05/2023	26/05/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	26/05/2023	26/05/2023
					CT314	Electrically powered	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	26/05/2023	26/05/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	26/05/2023	26/05/2023
					CT321	Long-term surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	26/05/2023	26/05/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	26/05/2023	26/05/2023
					CT449	Generators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	26/05/2023	26/05/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	26/05/2023	26/05/2023
					CT981	Single use	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	26/05/2023	26/05/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	26/05/2023	26/05/2023
					CT983	Surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	26/05/2023	26/05/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	26/05/2023	26/05/2023
13476	2023	All other reportable incidents	17882	Physiologic-monitoring defibrillation system	CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A07	Electrical /Electronic Property Problem		
							F27	Problem Identified During Non-Clinical Procedure	A07	Electrical /Electronic Property Problem		
13477	2023	All other reportable incidents	35852	Single-chamber implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
13478	2023	All other reportable incidents	47265	Dual-chamber implantable pacemaker, rate-responsive	CT141	Electrodes/Leads and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070909	Oversensing		
									A090801	Signal Artifact/Noise		
							F08	Hospitalization or Prolonged Hospitalization	A070909	Oversensing		
									A090801	Signal Artifact/Noise		
							F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A070909	Oversensing		
									A090801	Signal Artifact/Noise		
							F1905	Device Revision or Replacement	A070909	Oversensing		
									A090801	Signal Artifact/Noise		
					CT145	Cardiovascular devices	E2403	No Clinical Signs, Symptoms or Conditions	A070909	Oversensing		
									A090801	Signal Artifact/Noise		
							F08	Hospitalization or Prolonged Hospitalization	A070909	Oversensing		
									A090801	Signal Artifact/Noise		
							F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A070909	Oversensing		
									A090801	Signal Artifact/Noise		
							F1905	Device Revision or Replacement	A070909	Oversensing		
									A090801	Signal Artifact/Noise		
					CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070909	Oversensing		
									A090801	Signal Artifact/Noise		
							F08	Hospitalization or Prolonged Hospitalization	A070909	Oversensing		
									A090801	Signal Artifact/Noise		
							F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A070909	Oversensing		
									A090801	Signal Artifact/Noise		
							F1905	Device Revision or Replacement	A070909	Oversensing		
									A090801	Signal Artifact/Noise		
					CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A070909	Oversensing		
									A090801	Signal Artifact/Noise		
							F08	Hospitalization or Prolonged Hospitalization	A070909	Oversensing		
									A090801	Signal Artifact/Noise		
							F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A070909	Oversensing		
									A090801	Signal Artifact/Noise		
							F1905	Device Revision or Replacement	A070909	Oversensing		
									A090801	Signal Artifact/Noise		
					CT233	Surgical	E2403	No Clinical Signs, Symptoms or Conditions	A070909	Oversensing		
									A090801	Signal Artifact/Noise		
							F08	Hospitalization or Prolonged Hospitalization	A070909	Oversensing		
									A090801	Signal Artifact/Noise		
							F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A070909	Oversensing		
									A090801	Signal Artifact/Noise		
							F1905	Device Revision or Replacement	A070909	Oversensing		
									A090801	Signal Artifact/Noise		
					CT2406	Implantable	E2403	No Clinical Signs, Symptoms or Conditions	A070909	Oversensing		
									A090801	Signal Artifact/Noise		
							F08	Hospitalization or Prolonged Hospitalization	A070909	Oversensing		
									A090801	Signal Artifact/Noise		
							F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A070909	Oversensing		
									A090801	Signal Artifact/Noise		
							F1905	Device Revision or Replacement	A070909	Oversensing		
									A090801	Signal Artifact/Noise		
					CT2407	Active implantable	E2403	No Clinical Signs, Symptoms or Conditions	A070909	Oversensing		
									A090801	Signal Artifact/Noise		
							F08	Hospitalization or Prolonged Hospitalization	A070909	Oversensing		
									A090801	Signal Artifact/Noise		
							F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A070909	Oversensing		
									A090801	Signal Artifact/Noise		
							F1905	Device Revision or Replacement	A070909	Oversensing		
									A090801	Signal Artifact/Noise		
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A070909	Oversensing		
									A090801	Signal Artifact/Noise		
							F08	Hospitalization or Prolonged Hospitalization	A070909	Oversensing		
									A090801	Signal Artifact/Noise		
							F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A070909	Oversensing		
									A090801	Signal Artifact/Noise		
							F1905	Device Revision or Replacement	A070909	Oversensing		

									A090801	Signal Artifact/Noise		
					CT2727	Active implant devices/ancillaries	E2403	No Clinical Signs, Symptoms or Conditions	A070909	Oversensing		
									A090801	Signal Artifact/Noise		
							F08	Hospitalization or Prolonged Hospitalization	A070909	Oversensing		
									A090801	Signal Artifact/Noise		
							F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A070909	Oversensing		
									A090801	Signal Artifact/Noise		
							F1905	Device Revision or Replacement	A070909	Oversensing		
									A090801	Signal Artifact/Noise		
					CT2750	Implantable cardiac electrical stimulators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070909	Oversensing		
									A090801	Signal Artifact/Noise		
							F08	Hospitalization or Prolonged Hospitalization	A070909	Oversensing		
									A090801	Signal Artifact/Noise		
							F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A070909	Oversensing		
									A090801	Signal Artifact/Noise		
							F1905	Device Revision or Replacement	A070909	Oversensing		
									A090801	Signal Artifact/Noise		
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A070909	Oversensing		
									A090801	Signal Artifact/Noise		
							F08	Hospitalization or Prolonged Hospitalization	A070909	Oversensing		
									A090801	Signal Artifact/Noise		
							F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A070909	Oversensing		
									A090801	Signal Artifact/Noise		
							F1905	Device Revision or Replacement	A070909	Oversensing		
									A090801	Signal Artifact/Noise		
					CT2985	Active	E2403	No Clinical Signs, Symptoms or Conditions	A070909	Oversensing		
									A090801	Signal Artifact/Noise		
							F08	Hospitalization or Prolonged Hospitalization	A070909	Oversensing		
									A090801	Signal Artifact/Noise		
							F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A070909	Oversensing		
									A090801	Signal Artifact/Noise		
							F1905	Device Revision or Replacement	A070909	Oversensing		
									A090801	Signal Artifact/Noise		
					CT3001	Implantable metal	E2403	No Clinical Signs, Symptoms or Conditions	A070909	Oversensing		
									A090801	Signal Artifact/Noise		
							F08	Hospitalization or Prolonged Hospitalization	A070909	Oversensing		
									A090801	Signal Artifact/Noise		
							F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A070909	Oversensing		
									A090801	Signal Artifact/Noise		
							F1905	Device Revision or Replacement	A070909	Oversensing		
									A090801	Signal Artifact/Noise		
					CT3002	Implantable synthetic polymer	E2403	No Clinical Signs, Symptoms or Conditions	A070909	Oversensing		
									A090801	Signal Artifact/Noise		
							F08	Hospitalization or Prolonged Hospitalization	A070909	Oversensing		
									A090801	Signal Artifact/Noise		
							F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A070909	Oversensing		
									A090801	Signal Artifact/Noise		
							F1905	Device Revision or Replacement	A070909	Oversensing		
									A090801	Signal Artifact/Noise		
					CT3006	Implantable inorganic material	E2403	No Clinical Signs, Symptoms or Conditions	A070909	Oversensing		
									A090801	Signal Artifact/Noise		
							F08	Hospitalization or Prolonged Hospitalization	A070909	Oversensing		
									A090801	Signal Artifact/Noise		
							F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A070909	Oversensing		
									A090801	Signal Artifact/Noise		
							F1905	Device Revision or Replacement	A070909	Oversensing		
									A090801	Signal Artifact/Noise		
					CT314	Electrically powered	E2403	No Clinical Signs, Symptoms or Conditions	A070909	Oversensing		
									A090801	Signal Artifact/Noise		
							F08	Hospitalization or Prolonged Hospitalization	A070909	Oversensing		
									A090801	Signal Artifact/Noise		
							F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A070909	Oversensing		
									A090801	Signal Artifact/Noise		
							F1905	Device Revision or Replacement	A070909	Oversensing		
									A090801	Signal Artifact/Noise		
					CT321	Long-term surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A070909	Oversensing		
									A090801	Signal Artifact/Noise		
							F08	Hospitalization or Prolonged Hospitalization	A070909	Oversensing		
									A090801	Signal Artifact/Noise		
							F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A070909	Oversensing		
									A090801	Signal Artifact/Noise		
							F1905	Device Revision or Replacement	A070909	Oversensing		
									A090801	Signal Artifact/Noise		
					CT449	Generators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070909	Oversensing		
									A090801	Signal Artifact/Noise		
							F08	Hospitalization or Prolonged Hospitalization	A070909	Oversensing		
									A090801	Signal Artifact/Noise		
							F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A070909	Oversensing		
									A090801	Signal Artifact/Noise		
							F1905	Device Revision or Replacement	A070909	Oversensing		
									A090801	Signal Artifact/Noise		
					CT981	Single use	E2403	No Clinical Signs, Symptoms or Conditions	A070909	Oversensing		
									A090801	Signal Artifact/Noise		
							F08	Hospitalization or Prolonged Hospitalization	A070909	Oversensing		
									A090801	Signal Artifact/Noise		
							F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A070909	Oversensing		
									A090801	Signal Artifact/Noise		
							F1905	Device Revision or Replacement	A070909	Oversensing		

									A090801	Signal Artifact/Noise		
					CT983	Surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A070909	Oversensing		
									A090801	Signal Artifact/Noise		
							F08	Hospitalization or Prolonged Hospitalization	A070909	Oversensing		
									A090801	Signal Artifact/Noise		
							F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A070909	Oversensing		
									A090801	Signal Artifact/Noise		
							F1905	Device Revision or Replacement	A070909	Oversensing		
									A090801	Signal Artifact/Noise		
13479	2023	All other reportable incidents	17882	Physiologic-monitoring defibrillation system	CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070803	Failure to Power Up		
							F27	Problem Identified During Non-Clinical Procedure	A070803	Failure to Power Up		
13480	2023	All other reportable incidents	35223	Endocardial/interventricular septal pacing lead	CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A051201	Device Dislodged or Dislocated		
									A070101	Failure to Capture		
									A070909	Oversensing		
							F08	Hospitalization or Prolonged Hospitalization	A051201	Device Dislodged or Dislocated		
									A070101	Failure to Capture		
									A070909	Oversensing		
							F12	Serious Injury/ Illness/ Impairment	A051201	Device Dislodged or Dislocated		
									A070101	Failure to Capture		
									A070909	Oversensing		
							F1904	Device Repositioning	A051201	Device Dislodged or Dislocated		
									A070101	Failure to Capture		
									A070909	Oversensing		
13481	2023	All other reportable incidents	17882	Physiologic-monitoring defibrillation system	CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A0801	Failure to Calibrate		
							F27	Problem Identified During Non-Clinical Procedure	A0801	Failure to Calibrate		
13482	2023	All other reportable incidents	35852	Single-chamber implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E2104	Electric Shock	A070909	Oversensing		
									A071302	Inappropriate/Inadequate Shock/Stimulation		
									A090801	Signal Artifact/Noise		
									A1301	Failure to Read Input Signal		
							F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A070909	Oversensing		
									A071302	Inappropriate/Inadequate Shock/Stimulation		
									A090801	Signal Artifact/Noise		
									A1301	Failure to Read Input Signal		
							F2203	Imaging Required	A070909	Oversensing		
									A071302	Inappropriate/Inadequate Shock/Stimulation		
									A090801	Signal Artifact/Noise		
									A1301	Failure to Read Input Signal		
13483	2023	All other reportable incidents	47263	Cardiac resynchronization therapy implantable pacemaker	CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	27/06/2023	27/06/2023
							F1901	Additional Surgery	A070504	Premature Discharge of Battery	27/06/2023	27/06/2023
					CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	27/06/2023	27/06/2023
							F1901	Additional Surgery	A070504	Premature Discharge of Battery	27/06/2023	27/06/2023
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	27/06/2023	27/06/2023
							F1901	Additional Surgery	A070504	Premature Discharge of Battery	27/06/2023	27/06/2023
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	27/06/2023	27/06/2023
							F1901	Additional Surgery	A070504	Premature Discharge of Battery	27/06/2023	27/06/2023
13484	2023	All other reportable incidents	47265	Dual-chamber implantable pacemaker, rate-responsive	CT141	Electrodes/Leads and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	04/06/2023	04/06/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	04/06/2023	04/06/2023
					CT145	Cardiovascular devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	04/06/2023	04/06/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	04/06/2023	04/06/2023
					CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	04/06/2023	04/06/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	04/06/2023	04/06/2023
					CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	04/06/2023	04/06/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	04/06/2023	04/06/2023
					CT233	Surgical	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	04/06/2023	04/06/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	04/06/2023	04/06/2023
					CT2406	Implantable	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	04/06/2023	04/06/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	04/06/2023	04/06/2023
					CT2407	Active implantable	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	04/06/2023	04/06/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	04/06/2023	04/06/2023
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	04/06/2023	04/06/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	04/06/2023	04/06/2023
					CT2727	Active implant devices/ancillaries	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	04/06/2023	04/06/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	04/06/2023	04/06/2023
					CT2750	Implantable cardiac electrical stimulators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	04/06/2023	04/06/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	04/06/2023	04/06/2023
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	04/06/2023	04/06/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	04/06/2023	04/06/2023
					CT2985	Active	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	04/06/2023	04/06/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	04/06/2023	04/06/2023
					CT3001	Implantable metal	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	04/06/2023	04/06/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	04/06/2023	04/06/2023
					CT3002	Implantable synthetic polymer	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	04/06/2023	04/06/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	04/06/2023	04/06/2023
					CT3006	Implantable inorganic material	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	04/06/2023	04/06/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	04/06/2023	04/06/2023
					CT314	Electrically powered	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	04/06/2023	04/06/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	04/06/2023	04/06/2023
					CT321	Long-term surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	04/06/2023	04/06/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	04/06/2023	04/06/2023
					CT449	Generators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	04/06/2023	04/06/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	04/06/2023	04/06/2023
					CT981	Single use	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	04/06/2023	04/06/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	04/06/2023	04/06/2023
					CT983	Surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	04/06/2023	04/06/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	04/06/2023	04/06/2023
13485	2023		47265	Dual-chamber implantable pacemaker, rate-responsive	CT141	Electrodes/Leads and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	26/06/2023	26/06/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	26/06/2023	26/06/2023
					CT145	Cardiovascular devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	26/06/2023	26/06/2023

							F1901	Additional Surgery	A24	No Apparent Device Problem	26/06/2023	26/06/2023
					CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	26/06/2023	26/06/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	26/06/2023	26/06/2023
					CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	26/06/2023	26/06/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	26/06/2023	26/06/2023
					CT233	Surgical	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	26/06/2023	26/06/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	26/06/2023	26/06/2023
					CT2406	Implantable	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	26/06/2023	26/06/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	26/06/2023	26/06/2023
					CT2407	Active implantable	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	26/06/2023	26/06/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	26/06/2023	26/06/2023
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	26/06/2023	26/06/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	26/06/2023	26/06/2023
					CT2727	Active implant devices/ancillaries	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	26/06/2023	26/06/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	26/06/2023	26/06/2023
					CT2750	Implantable cardiac electrical stimulators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	26/06/2023	26/06/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	26/06/2023	26/06/2023
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	26/06/2023	26/06/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	26/06/2023	26/06/2023
					CT2985	Active	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	26/06/2023	26/06/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	26/06/2023	26/06/2023
					CT3001	Implantable metal	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	26/06/2023	26/06/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	26/06/2023	26/06/2023
					CT3002	Implantable synthetic polymer	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	26/06/2023	26/06/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	26/06/2023	26/06/2023
					CT3006	Implantable inorganic material	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	26/06/2023	26/06/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	26/06/2023	26/06/2023
					CT314	Electrically powered	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	26/06/2023	26/06/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	26/06/2023	26/06/2023
					CT321	Long-term surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	26/06/2023	26/06/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	26/06/2023	26/06/2023
					CT449	Generators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	26/06/2023	26/06/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	26/06/2023	26/06/2023
					CT981	Single use	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	26/06/2023	26/06/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	26/06/2023	26/06/2023
					CT983	Surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	26/06/2023	26/06/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	26/06/2023	26/06/2023
13486	2023	All other reportable incidents	17882	Physiologic-monitoring defibrillation system	CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A0719	Unexpected Shutdown		
							F27	Problem Identified During Non-Clinical Procedure	A0719	Unexpected Shutdown		
13487	2023	All other reportable incidents	17882	Physiologic-monitoring defibrillation system	CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A0719	Unexpected Shutdown		
							F26	No Health Consequences or Impact	A0719	Unexpected Shutdown		
13488	2023	Death	47267	Single-chamber implantable pacemaker, rate-responsive	CT141	Electrodes/Leads and associated devices	E060109	Tachycardia	A24	No Apparent Device Problem		
							E060110	Ventricular Fibrillation	A24	No Apparent Device Problem		
							F02	Death	A24	No Apparent Device Problem		
					CT145	Cardiovascular devices	E060109	Tachycardia	A24	No Apparent Device Problem		
							E060110	Ventricular Fibrillation	A24	No Apparent Device Problem		
							F02	Death	A24	No Apparent Device Problem		
					CT1684	Cardiac pacemakers and associated devices	E060109	Tachycardia	A24	No Apparent Device Problem		
							E060110	Ventricular Fibrillation	A24	No Apparent Device Problem		
							F02	Death	A24	No Apparent Device Problem		
					CT1685	Cardiac pacemakers	E060109	Tachycardia	A24	No Apparent Device Problem		
							E060110	Ventricular Fibrillation	A24	No Apparent Device Problem		
							F02	Death	A24	No Apparent Device Problem		
					CT233	Surgical	E060109	Tachycardia	A24	No Apparent Device Problem		
							E060110	Ventricular Fibrillation	A24	No Apparent Device Problem		
							F02	Death	A24	No Apparent Device Problem		
					CT2406	Implantable	E060109	Tachycardia	A24	No Apparent Device Problem		
							E060110	Ventricular Fibrillation	A24	No Apparent Device Problem		
							F02	Death	A24	No Apparent Device Problem		
					CT2407	Active implantable	E060109	Tachycardia	A24	No Apparent Device Problem		
							E060110	Ventricular Fibrillation	A24	No Apparent Device Problem		
							F02	Death	A24	No Apparent Device Problem		
					CT257	Implantable cardiac pacemakers	E060109	Tachycardia	A24	No Apparent Device Problem		
							E060110	Ventricular Fibrillation	A24	No Apparent Device Problem		
							F02	Death	A24	No Apparent Device Problem		
					CT2727	Active implant devices/ancillaries	E060109	Tachycardia	A24	No Apparent Device Problem		
							E060110	Ventricular Fibrillation	A24	No Apparent Device Problem		
							F02	Death	A24	No Apparent Device Problem		
					CT2750	Implantable cardiac electrical stimulators and associated devices	E060109	Tachycardia	A24	No Apparent Device Problem		
							E060110	Ventricular Fibrillation	A24	No Apparent Device Problem		
							F02	Death	A24	No Apparent Device Problem		
					CT2866	Implantable cardiac electrical stimulators	E060109	Tachycardia	A24	No Apparent Device Problem		
							E060110	Ventricular Fibrillation	A24	No Apparent Device Problem		
							F02	Death	A24	No Apparent Device Problem		
					CT2985	Active	E060109	Tachycardia	A24	No Apparent Device Problem		
							E060110	Ventricular Fibrillation	A24	No Apparent Device Problem		
							F02	Death	A24	No Apparent Device Problem		
					CT3001	Implantable metal	E060109	Tachycardia	A24	No Apparent Device Problem		
							E060110	Ventricular Fibrillation	A24	No Apparent Device Problem		
							F02	Death	A24	No Apparent Device Problem		
					CT3002	Implantable synthetic polymer	E060109	Tachycardia	A24	No Apparent Device Problem		
							E060110	Ventricular Fibrillation	A24	No Apparent Device Problem		
							F02	Death	A24	No Apparent Device Problem		
					CT3006	Implantable inorganic material	E060109	Tachycardia	A24	No Apparent Device Problem		
							E060110	Ventricular Fibrillation	A24	No Apparent Device Problem		
							F02	Death	A24	No Apparent Device Problem		
					CT314	Electrically powered	E060109	Tachycardia	A24	No Apparent Device Problem		
							E060110	Ventricular Fibrillation	A24	No Apparent Device Problem		

							F02	Death	A24	No Apparent Device Problem		
					CT321	Long-term surgical invasive	E060109	Tachycardia	A24	No Apparent Device Problem		
							E060110	Ventricular Fibrillation	A24	No Apparent Device Problem		
							F02	Death	A24	No Apparent Device Problem		
					CT449	Generators and associated devices	E060109	Tachycardia	A24	No Apparent Device Problem		
							E060110	Ventricular Fibrillation	A24	No Apparent Device Problem		
							F02	Death	A24	No Apparent Device Problem		
					CT981	Single use	E060109	Tachycardia	A24	No Apparent Device Problem		
							E060110	Ventricular Fibrillation	A24	No Apparent Device Problem		
							F02	Death	A24	No Apparent Device Problem		
					CT983	Surgical invasive	E060109	Tachycardia	A24	No Apparent Device Problem		
							E060110	Ventricular Fibrillation	A24	No Apparent Device Problem		
							F02	Death	A24	No Apparent Device Problem		
13489	2023	All other reportable incidents	35223	Endocardial/interventricular septal pacing lead	CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A0722	Impedance Problem		
							F26	No Health Consequences or Impact	A0722	Impedance Problem		
13490	2023	All other reportable incidents	47263	Cardiac resynchronization therapy implantable pacemaker	CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem		
									A070909	Oversensing		
									A090801	Signal Artifact/Noise		
									A1301	Failure to Read Input Signal		
							F26	No Health Consequences or Impact	A0705	Battery Problem		
									A070909	Oversensing		
									A090801	Signal Artifact/Noise		
									A1301	Failure to Read Input Signal		
					CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem		
									A070909	Oversensing		
									A090801	Signal Artifact/Noise		
									A1301	Failure to Read Input Signal		
							F26	No Health Consequences or Impact	A0705	Battery Problem		
									A070909	Oversensing		
									A090801	Signal Artifact/Noise		
									A1301	Failure to Read Input Signal		
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem		
									A070909	Oversensing		
									A090801	Signal Artifact/Noise		
									A1301	Failure to Read Input Signal		
							F26	No Health Consequences or Impact	A0705	Battery Problem		
									A070909	Oversensing		
									A090801	Signal Artifact/Noise		
									A1301	Failure to Read Input Signal		
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem		
									A070909	Oversensing		
									A090801	Signal Artifact/Noise		
									A1301	Failure to Read Input Signal		
							F26	No Health Consequences or Impact	A0705	Battery Problem		
									A070909	Oversensing		
									A090801	Signal Artifact/Noise		
									A1301	Failure to Read Input Signal		
13491	2023	All other reportable incidents	35852	Single-chamber implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A12	Connection Problem	21/04/2016	21/04/2016
							F1908	Prolonged Surgery	A12	Connection Problem	21/04/2016	21/04/2016
13492	2023	All other reportable incidents	47265	Dual-chamber implantable pacemaker, rate-responsive	CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem	14/07/2023	14/07/2023
									A070504	Premature Discharge of Battery	14/07/2023	14/07/2023
							F08	Hospitalization or Prolonged Hospitalization	A0705	Battery Problem	14/07/2023	14/07/2023
									A070504	Premature Discharge of Battery	14/07/2023	14/07/2023
							F1905	Device Revision or Replacement	A0705	Battery Problem	14/07/2023	14/07/2023
									A070504	Premature Discharge of Battery	14/07/2023	14/07/2023
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem	14/07/2023	14/07/2023
									A070504	Premature Discharge of Battery	14/07/2023	14/07/2023
							F08	Hospitalization or Prolonged Hospitalization	A0705	Battery Problem	14/07/2023	14/07/2023
									A070504	Premature Discharge of Battery	14/07/2023	14/07/2023
							F1905	Device Revision or Replacement	A0705	Battery Problem	14/07/2023	14/07/2023
									A070504	Premature Discharge of Battery	14/07/2023	14/07/2023
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem	14/07/2023	14/07/2023
									A070504	Premature Discharge of Battery	14/07/2023	14/07/2023
							F08	Hospitalization or Prolonged Hospitalization	A0705	Battery Problem	14/07/2023	14/07/2023
									A070504	Premature Discharge of Battery	14/07/2023	14/07/2023
							F1905	Device Revision or Replacement	A0705	Battery Problem	14/07/2023	14/07/2023
									A070504	Premature Discharge of Battery	14/07/2023	14/07/2023
13493	2023	All other reportable incidents	47263	Cardiac resynchronization therapy implantable pacemaker	CT1684	Cardiac pacemakers and associated devices	E0717	Dyspnea	A090804	False Positive Result		
							F08	Hospitalization or Prolonged Hospitalization	A090804	False Positive Result		
							F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A090804	False Positive Result		
					CT1685	Cardiac pacemakers	E0717	Dyspnea	A090804	False Positive Result		
							F08	Hospitalization or Prolonged Hospitalization	A090804	False Positive Result		
							F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A090804	False Positive Result		
					CT257	Implantable cardiac pacemakers	E0717	Dyspnea	A090804	False Positive Result		
							F08	Hospitalization or Prolonged Hospitalization	A090804	False Positive Result		
							F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A090804	False Positive Result		
					CT2866	Implantable cardiac electrical stimulators	E0717	Dyspnea	A090804	False Positive Result		
							F08	Hospitalization or Prolonged Hospitalization	A090804	False Positive Result		
							F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A090804	False Positive Result		
13494	2023	All other reportable incidents	35223	Endocardial/interventricular septal pacing lead	CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A051201	Device Dislodged or Dislocated		
									A070909	Oversensing		
							F08	Hospitalization or Prolonged Hospitalization	A051201	Device Dislodged or Dislocated		
									A070909	Oversensing		
							F12	Serious Injury/ Illness/ Impairment	A051201	Device Dislodged or Dislocated		
									A070909	Oversensing		
							F1904	Device Repositioning	A051201	Device Dislodged or Dislocated		
									A070909	Oversensing		
13495	2023	All other reportable incidents	47270	Cardiac resynchronization therapy implantable	CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	25/07/2023	25/07/2023

				defibrillator					A0908	Incorrect, Inadequate or Imprecise Result or Readings	25/07/2023	25/07/2023
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	25/07/2023	25/07/2023
									A0908	Incorrect, Inadequate or Imprecise Result or Readings	25/07/2023	25/07/2023
							F1903	Device Explantation	A070504	Premature Discharge of Battery	25/07/2023	25/07/2023
									A0908	Incorrect, Inadequate or Imprecise Result or Readings	25/07/2023	25/07/2023
13496	2023	All other reportable incidents	37265	Dual-chamber implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E2104	Electric Shock	A070909	Oversensing		
									A071302	Inappropriate/Inadequate Shock/Stimulation		
							F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A070909	Oversensing		
									A071302	Inappropriate/Inadequate Shock/Stimulation		
13497	2023	All other reportable incidents	35852	Single-chamber implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	03/02/2023	03/02/2023
									A0908	Incorrect, Inadequate or Imprecise Result or Readings	03/02/2023	03/02/2023
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	03/02/2023	03/02/2023
									A0908	Incorrect, Inadequate or Imprecise Result or Readings	03/02/2023	03/02/2023
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	03/02/2023	03/02/2023
									A0908	Incorrect, Inadequate or Imprecise Result or Readings	03/02/2023	03/02/2023
13498	2023	All other reportable incidents	35223	Endocardial/interventricular septal pacing lead	CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A072201	High Impedance		
							F12	Serious Injury/ Illness/ Impairment	A072201	High Impedance		
							F23	Unexpected Medical Intervention	A072201	High Impedance		
							F26	No Health Consequences or Impact	A072201	High Impedance		
13499	2023	All other reportable incidents	35852	Single-chamber implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A0508	Noise, Audible	27/06/2023	27/06/2023
									A0705	Battery Problem	27/06/2023	27/06/2023
									A070504	Premature Discharge of Battery	27/06/2023	27/06/2023
									A1304	Telemetry Discrepancy	27/06/2023	27/06/2023
							F08	Hospitalization or Prolonged Hospitalization	A0508	Noise, Audible	27/06/2023	27/06/2023
									A0705	Battery Problem	27/06/2023	27/06/2023
									A070504	Premature Discharge of Battery	27/06/2023	27/06/2023
									A1304	Telemetry Discrepancy	27/06/2023	27/06/2023
							F1905	Device Revision or Replacement	A0508	Noise, Audible	27/06/2023	27/06/2023
									A0705	Battery Problem	27/06/2023	27/06/2023
									A070504	Premature Discharge of Battery	27/06/2023	27/06/2023
									A1304	Telemetry Discrepancy	27/06/2023	27/06/2023
13500	2023	All other reportable incidents	17882	Physiologic-monitoring defibrillation system	CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A07	Electrical /Electronic Property Problem		
							F27	Problem Identified During Non-Clinical Procedure	A07	Electrical /Electronic Property Problem		
13501	2023	All other reportable incidents	47263	Cardiac resynchronization therapy implantable pacemaker	CT1685	Cardiac pacemakers	E0602	Cardiac Arrest	A070101	Failure to Capture	17/06/2023	17/06/2023
									A0908	Incorrect, Inadequate or Imprecise Result or Readings	17/06/2023	17/06/2023
							F08	Hospitalization or Prolonged Hospitalization	A070101	Failure to Capture	17/06/2023	17/06/2023
									A0908	Incorrect, Inadequate or Imprecise Result or Readings	17/06/2023	17/06/2023
							F1006	Inadequate Treatment/Disease Progression	A070101	Failure to Capture	17/06/2023	17/06/2023
									A0908	Incorrect, Inadequate or Imprecise Result or Readings	17/06/2023	17/06/2023
							F1905	Device Revision or Replacement	A070101	Failure to Capture	17/06/2023	17/06/2023
									A0908	Incorrect, Inadequate or Imprecise Result or Readings	17/06/2023	17/06/2023
							F2301	Additional Device Required	A070101	Failure to Capture	17/06/2023	17/06/2023
									A0908	Incorrect, Inadequate or Imprecise Result or Readings	17/06/2023	17/06/2023
					CT257	Implantable cardiac pacemakers	E0602	Cardiac Arrest	A070101	Failure to Capture	17/06/2023	17/06/2023
									A0908	Incorrect, Inadequate or Imprecise Result or Readings	17/06/2023	17/06/2023
							F08	Hospitalization or Prolonged Hospitalization	A070101	Failure to Capture	17/06/2023	17/06/2023
									A0908	Incorrect, Inadequate or Imprecise Result or Readings	17/06/2023	17/06/2023
							F1006	Inadequate Treatment/Disease Progression	A070101	Failure to Capture	17/06/2023	17/06/2023
									A0908	Incorrect, Inadequate or Imprecise Result or Readings	17/06/2023	17/06/2023
							F1905	Device Revision or Replacement	A070101	Failure to Capture	17/06/2023	17/06/2023
									A0908	Incorrect, Inadequate or Imprecise Result or Readings	17/06/2023	17/06/2023
							F2301	Additional Device Required	A070101	Failure to Capture	17/06/2023	17/06/2023
									A0908	Incorrect, Inadequate or Imprecise Result or Readings	17/06/2023	17/06/2023
					CT2866	Implantable cardiac electrical stimulators	E0602	Cardiac Arrest	A070101	Failure to Capture	17/06/2023	17/06/2023
									A0908	Incorrect, Inadequate or Imprecise Result or Readings	17/06/2023	17/06/2023
							F08	Hospitalization or Prolonged Hospitalization	A070101	Failure to Capture	17/06/2023	17/06/2023
									A0908	Incorrect, Inadequate or Imprecise Result or Readings	17/06/2023	17/06/2023
							F1006	Inadequate Treatment/Disease Progression	A070101	Failure to Capture	17/06/2023	17/06/2023
									A0908	Incorrect, Inadequate or Imprecise Result or Readings	17/06/2023	17/06/2023
							F1905	Device Revision or Replacement	A070101	Failure to Capture	17/06/2023	17/06/2023
									A0908	Incorrect, Inadequate or Imprecise Result or Readings	17/06/2023	17/06/2023
							F2301	Additional Device Required	A070101	Failure to Capture	17/06/2023	17/06/2023
									A0908	Incorrect, Inadequate or Imprecise Result or Readings	17/06/2023	17/06/2023
13502	2023	All other reportable incidents	47263	Cardiac resynchronization therapy implantable pacemaker	CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A160601	Premature Elective Replacement Indicator		
							F1903	Device Explantation	A160601	Premature Elective Replacement Indicator		
					CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A160601	Premature Elective Replacement Indicator		
							F1903	Device Explantation	A160601	Premature Elective Replacement Indicator		
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A160601	Premature Elective Replacement Indicator		
							F1903	Device Explantation	A160601	Premature Elective Replacement Indicator		
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A160601	Premature Elective Replacement Indicator		
							F1903	Device Explantation	A160601	Premature Elective Replacement Indicator		
13503	2023	All other reportable incidents	47265	Dual-chamber implantable pacemaker, rate-	CT141	Electrodes/Leads and associated devices	E0109	Convulsion/Seizure	A05	Mechanical Problem		

				responsive					A1205	Loose or Intermittent Connection		
							F12	Serious Injury/ Illness/ Impairment	A05	Mechanical Problem		
									A1205	Loose or Intermittent Connection		
					CT145	Cardiovascular devices	E0109	Convulsion/Seizure	A05	Mechanical Problem		
									A1205	Loose or Intermittent Connection		
							F12	Serious Injury/ Illness/ Impairment	A05	Mechanical Problem		
									A1205	Loose or Intermittent Connection		
					CT1684	Cardiac pacemakers and associated devices	E0109	Convulsion/Seizure	A05	Mechanical Problem		
									A1205	Loose or Intermittent Connection		
							F12	Serious Injury/ Illness/ Impairment	A05	Mechanical Problem		
									A1205	Loose or Intermittent Connection		
					CT1685	Cardiac pacemakers	E0109	Convulsion/Seizure	A05	Mechanical Problem		
									A1205	Loose or Intermittent Connection		
							F12	Serious Injury/ Illness/ Impairment	A05	Mechanical Problem		
									A1205	Loose or Intermittent Connection		
					CT233	Surgical	E0109	Convulsion/Seizure	A05	Mechanical Problem		
									A1205	Loose or Intermittent Connection		
							F12	Serious Injury/ Illness/ Impairment	A05	Mechanical Problem		
									A1205	Loose or Intermittent Connection		
					CT2406	Implantable	E0109	Convulsion/Seizure	A05	Mechanical Problem		
									A1205	Loose or Intermittent Connection		
							F12	Serious Injury/ Illness/ Impairment	A05	Mechanical Problem		
									A1205	Loose or Intermittent Connection		
					CT2407	Active implantable	E0109	Convulsion/Seizure	A05	Mechanical Problem		
									A1205	Loose or Intermittent Connection		
							F12	Serious Injury/ Illness/ Impairment	A05	Mechanical Problem		
									A1205	Loose or Intermittent Connection		
					CT257	Implantable cardiac pacemakers	E0109	Convulsion/Seizure	A05	Mechanical Problem		
									A1205	Loose or Intermittent Connection		
							F12	Serious Injury/ Illness/ Impairment	A05	Mechanical Problem		
									A1205	Loose or Intermittent Connection		
					CT2727	Active implant devices/ancillaries	E0109	Convulsion/Seizure	A05	Mechanical Problem		
									A1205	Loose or Intermittent Connection		
							F12	Serious Injury/ Illness/ Impairment	A05	Mechanical Problem		
									A1205	Loose or Intermittent Connection		
					CT2750	Implantable cardiac electrical stimulators and associated devices	E0109	Convulsion/Seizure	A05	Mechanical Problem		
									A1205	Loose or Intermittent Connection		
							F12	Serious Injury/ Illness/ Impairment	A05	Mechanical Problem		
									A1205	Loose or Intermittent Connection		
					CT2866	Implantable cardiac electrical stimulators	E0109	Convulsion/Seizure	A05	Mechanical Problem		
									A1205	Loose or Intermittent Connection		
							F12	Serious Injury/ Illness/ Impairment	A05	Mechanical Problem		
									A1205	Loose or Intermittent Connection		
					CT2985	Active	E0109	Convulsion/Seizure	A05	Mechanical Problem		
									A1205	Loose or Intermittent Connection		
							F12	Serious Injury/ Illness/ Impairment	A05	Mechanical Problem		
									A1205	Loose or Intermittent Connection		
					CT3001	Implantable metal	E0109	Convulsion/Seizure	A05	Mechanical Problem		
									A1205	Loose or Intermittent Connection		
							F12	Serious Injury/ Illness/ Impairment	A05	Mechanical Problem		
									A1205	Loose or Intermittent Connection		
					CT3002	Implantable synthetic polymer	E0109	Convulsion/Seizure	A05	Mechanical Problem		
									A1205	Loose or Intermittent Connection		
							F12	Serious Injury/ Illness/ Impairment	A05	Mechanical Problem		
									A1205	Loose or Intermittent Connection		
					CT3006	Implantable inorganic material	E0109	Convulsion/Seizure	A05	Mechanical Problem		
									A1205	Loose or Intermittent Connection		
							F12	Serious Injury/ Illness/ Impairment	A05	Mechanical Problem		
									A1205	Loose or Intermittent Connection		
					CT314	Electrically powered	E0109	Convulsion/Seizure	A05	Mechanical Problem		
									A1205	Loose or Intermittent Connection		
							F12	Serious Injury/ Illness/ Impairment	A05	Mechanical Problem		
									A1205	Loose or Intermittent Connection		
					CT321	Long-term surgical invasive	E0109	Convulsion/Seizure	A05	Mechanical Problem		
									A1205	Loose or Intermittent Connection		
							F12	Serious Injury/ Illness/ Impairment	A05	Mechanical Problem		
									A1205	Loose or Intermittent Connection		
					CT449	Generators and associated devices	E0109	Convulsion/Seizure	A05	Mechanical Problem		
									A1205	Loose or Intermittent Connection		
							F12	Serious Injury/ Illness/ Impairment	A05	Mechanical Problem		
									A1205	Loose or Intermittent Connection		
					CT981	Single use	E0109	Convulsion/Seizure	A05	Mechanical Problem		
									A1205	Loose or Intermittent Connection		
							F12	Serious Injury/ Illness/ Impairment	A05	Mechanical Problem		
									A1205	Loose or Intermittent Connection		
					CT983	Surgical invasive	E0109	Convulsion/Seizure	A05	Mechanical Problem		
									A1205	Loose or Intermittent Connection		
							F12	Serious Injury/ Illness/ Impairment	A05	Mechanical Problem		
									A1205	Loose or Intermittent Connection		
13504	2023	All other reportable incidents	47265	Dual-chamber implantable pacemaker, rate-responsive	CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	08/08/2023	08/08/2023
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	08/08/2023	08/08/2023
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	08/08/2023	08/08/2023
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	08/08/2023	08/08/2023
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	08/08/2023	08/08/2023
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	08/08/2023	08/08/2023
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	08/08/2023	08/08/2023
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	08/08/2023	08/08/2023
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	08/08/2023	08/08/2023

13505	2023	All other reportable incidents	47265	Dual-chamber implantable pacemaker, rate-responsive	CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	01/11/2023	01/11/2023
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	01/11/2023	01/11/2023
							F1903	Device Explantation	A070504	Premature Discharge of Battery	01/11/2023	01/11/2023
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	01/11/2023	01/11/2023
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	01/11/2023	01/11/2023
							F1903	Device Explantation	A070504	Premature Discharge of Battery	01/11/2023	01/11/2023
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	01/11/2023	01/11/2023
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	01/11/2023	01/11/2023
							F1903	Device Explantation	A070504	Premature Discharge of Battery	01/11/2023	01/11/2023
13506	2023	All other reportable incidents	63188	Left-ventricular wireless pacing system sensor/transmitter	CT141	Electrodes/Leads and associated devices			A070801	Complete Loss of Power		
					CT145	Cardiovascular devices			A070801	Complete Loss of Power		
					CT1577	Wireless monitoring system transmitters			A070801	Complete Loss of Power		
					CT1684	Cardiac pacemakers and associated devices			A070801	Complete Loss of Power		
					CT2317	Transmitters/Receivers/Transceivers			A070801	Complete Loss of Power		
					CT233	Surgical			A070801	Complete Loss of Power		
					CT2406	Implantable			A070801	Complete Loss of Power		
					CT2407	Active implantable			A070801	Complete Loss of Power		
					CT255	Transducers and associated devices			A070801	Complete Loss of Power		
					CT2726	Physiological monitoring/measuring devices			A070801	Complete Loss of Power		
					CT2727	Active implant devices/ancillaries			A070801	Complete Loss of Power		
					CT2750	Implantable cardiac electrical stimulators and associated devices			A070801	Complete Loss of Power		
					CT2885	Physiological monitoring/measuring system transmitters/receivers/transceivers			A070801	Complete Loss of Power		
					CT2985	Active			A070801	Complete Loss of Power		
					CT3001	Implantable metal			A070801	Complete Loss of Power		
					CT3002	Implantable synthetic polymer			A070801	Complete Loss of Power		
					CT3006	Implantable inorganic material			A070801	Complete Loss of Power		
					CT314	Electrically powered			A070801	Complete Loss of Power		
					CT321	Long-term surgical invasive			A070801	Complete Loss of Power		
					CT384	Sensors and associated devices			A070801	Complete Loss of Power		
					CT449	Generators and associated devices			A070801	Complete Loss of Power		
					CT981	Single use			A070801	Complete Loss of Power		
					CT983	Surgical invasive			A070801	Complete Loss of Power		
					CT994	Sensors			A070801	Complete Loss of Power		
13507	2023	All other reportable incidents	47263	Cardiac resynchronization therapy implantable pacemaker	CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A0908	Incorrect, Inadequate or Imprecise Result or Readings	21/06/2023	21/06/2023
							F08	Hospitalization or Prolonged Hospitalization	A0908	Incorrect, Inadequate or Imprecise Result or Readings	21/06/2023	21/06/2023
							F1905	Device Revision or Replacement	A0908	Incorrect, Inadequate or Imprecise Result or Readings	21/06/2023	21/06/2023
					CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A0908	Incorrect, Inadequate or Imprecise Result or Readings	21/06/2023	21/06/2023
							F08	Hospitalization or Prolonged Hospitalization	A0908	Incorrect, Inadequate or Imprecise Result or Readings	21/06/2023	21/06/2023
							F1905	Device Revision or Replacement	A0908	Incorrect, Inadequate or Imprecise Result or Readings	21/06/2023	21/06/2023
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A0908	Incorrect, Inadequate or Imprecise Result or Readings	21/06/2023	21/06/2023
							F08	Hospitalization or Prolonged Hospitalization	A0908	Incorrect, Inadequate or Imprecise Result or Readings	21/06/2023	21/06/2023
							F1905	Device Revision or Replacement	A0908	Incorrect, Inadequate or Imprecise Result or Readings	21/06/2023	21/06/2023
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A0908	Incorrect, Inadequate or Imprecise Result or Readings	21/06/2023	21/06/2023
							F08	Hospitalization or Prolonged Hospitalization	A0908	Incorrect, Inadequate or Imprecise Result or Readings	21/06/2023	21/06/2023
							F1905	Device Revision or Replacement	A0908	Incorrect, Inadequate or Imprecise Result or Readings	21/06/2023	21/06/2023
13508	2023	All other reportable incidents	47265	Dual-chamber implantable pacemaker, rate-responsive	CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
13509	2023	All other reportable incidents	35039	Epicardial pacing lead	CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A040101	Fracture	17/05/2023	17/05/2023
									A070909	Oversensing	17/05/2023	17/05/2023
									A090801	Signal Artifact/Noise	17/05/2023	17/05/2023
							F12	Serious Injury/ Illness/ Impairment	A040101	Fracture	17/05/2023	17/05/2023
									A070909	Oversensing	17/05/2023	17/05/2023
									A090801	Signal Artifact/Noise	17/05/2023	17/05/2023
							F23	Unexpected Medical Intervention	A040101	Fracture	17/05/2023	17/05/2023
									A070909	Oversensing	17/05/2023	17/05/2023
									A090801	Signal Artifact/Noise	17/05/2023	17/05/2023
13510	2023	All other reportable incidents	47267	Single-chamber implantable pacemaker, rate-responsive	CT141	Electrodes/Leads and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
					CT145	Cardiovascular devices	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
					CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A0705	Battery Problem		

									A070504	Premature Discharge of Battery		
					CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
					CT233	Surgical	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
					CT2406	Implantable	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
					CT2407	Active implantable	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
					CT2727	Active implant devices/ancillaries	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
					CT2750	Implantable cardiac electrical stimulators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
					CT2985	Active	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
					CT3001	Implantable metal	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
					CT3002	Implantable synthetic polymer	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
					CT3006	Implantable inorganic material	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
					CT314	Electrically powered	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
					CT321	Long-term surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
					CT449	Generators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
					CT981	Single use	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
					CT983	Surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
13511	2023	All other reportable incidents	37265	Dual-chamber implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A0712	Pacing Problem		
							F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A0712	Pacing Problem		
13512	2023	All other reportable incidents	35852	Single-chamber implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A070909	Oversensing		
13513	2023	All other reportable incidents	47270	Cardiac resynchronization therapy implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	04/07/2023	04/07/2023
							F1901	Additional Surgery	A070504	Premature Discharge of Battery	04/07/2023	04/07/2023
13514	2023	All other reportable incidents	47265	Dual-chamber implantable pacemaker, rate-responsive	CT141	Electrodes/Leads and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070910	Undersensing		
									A0712	Pacing Problem		
									A1301	Failure to Read Input Signal		
							F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A070910	Undersensing		
									A0712	Pacing Problem		
									A1301	Failure to Read Input Signal		
					CT145	Cardiovascular devices	E2403	No Clinical Signs, Symptoms or Conditions	A070910	Undersensing		
									A0712	Pacing Problem		
									A1301	Failure to Read Input Signal		
							F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A070910	Undersensing		
									A0712	Pacing Problem		
									A1301	Failure to Read Input Signal		
					CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070910	Undersensing		
									A0712	Pacing Problem		

								A1301	Failure to Read Input Signal		
						F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A070910	Undersensing		
								A0712	Pacing Problem		
								A1301	Failure to Read Input Signal		
					CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A070910	Undersensing		
									A0712	Pacing Problem		
									A1301	Failure to Read Input Signal		
							F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A070910	Undersensing		
									A0712	Pacing Problem		
									A1301	Failure to Read Input Signal		
					CT233	Surgical	E2403	No Clinical Signs, Symptoms or Conditions	A070910	Undersensing		
									A0712	Pacing Problem		
									A1301	Failure to Read Input Signal		
							F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A070910	Undersensing		
									A0712	Pacing Problem		
									A1301	Failure to Read Input Signal		
					CT2406	Implantable	E2403	No Clinical Signs, Symptoms or Conditions	A070910	Undersensing		
									A0712	Pacing Problem		
									A1301	Failure to Read Input Signal		
							F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A070910	Undersensing		
									A0712	Pacing Problem		
									A1301	Failure to Read Input Signal		
					CT2407	Active implantable	E2403	No Clinical Signs, Symptoms or Conditions	A070910	Undersensing		
									A0712	Pacing Problem		
									A1301	Failure to Read Input Signal		
							F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A070910	Undersensing		
									A0712	Pacing Problem		
									A1301	Failure to Read Input Signal		
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A070910	Undersensing		
									A0712	Pacing Problem		
									A1301	Failure to Read Input Signal		
							F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A070910	Undersensing		
									A0712	Pacing Problem		
									A1301	Failure to Read Input Signal		
					CT2727	Active implant devices/ancillaries	E2403	No Clinical Signs, Symptoms or Conditions	A070910	Undersensing		
									A0712	Pacing Problem		
									A1301	Failure to Read Input Signal		
							F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A070910	Undersensing		
									A0712	Pacing Problem		
									A1301	Failure to Read Input Signal		
					CT2750	Implantable cardiac electrical stimulators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070910	Undersensing		
									A0712	Pacing Problem		
									A1301	Failure to Read Input Signal		
							F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A070910	Undersensing		
									A0712	Pacing Problem		
									A1301	Failure to Read Input Signal		
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A070910	Undersensing		
									A0712	Pacing Problem		
									A1301	Failure to Read Input Signal		
							F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A070910	Undersensing		
									A0712	Pacing Problem		
									A1301	Failure to Read Input Signal		
					CT2985	Active	E2403	No Clinical Signs, Symptoms or Conditions	A070910	Undersensing		
									A0712	Pacing Problem		
									A1301	Failure to Read Input Signal		
							F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A070910	Undersensing		
									A0712	Pacing Problem		
									A1301	Failure to Read Input Signal		
					CT3001	Implantable metal	E2403	No Clinical Signs, Symptoms or Conditions	A070910	Undersensing		
									A0712	Pacing Problem		
									A1301	Failure to Read Input Signal		
							F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A070910	Undersensing		
									A0712	Pacing Problem		
									A1301	Failure to Read Input Signal		
					CT3002	Implantable synthetic polymer	E2403	No Clinical Signs, Symptoms or Conditions	A070910	Undersensing		
									A0712	Pacing Problem		
									A1301	Failure to Read Input Signal		
							F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A070910	Undersensing		
									A0712	Pacing Problem		
									A1301	Failure to Read Input Signal		
					CT3006	Implantable inorganic material	E2403	No Clinical Signs, Symptoms or Conditions	A070910	Undersensing		
									A0712	Pacing Problem		
									A1301	Failure to Read Input Signal		
							F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A070910	Undersensing		
									A0712	Pacing Problem		
									A1301	Failure to Read Input Signal		
					CT314	Electrically powered	E2403	No Clinical Signs, Symptoms or Conditions	A070910	Undersensing		
									A0712	Pacing Problem		
									A1301	Failure to Read Input Signal		
							F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A070910	Undersensing		
									A0712	Pacing Problem		
									A1301	Failure to Read Input Signal		
					CT321	Long-term surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A070910	Undersensing		
									A0712	Pacing Problem		
									A1301	Failure to Read Input Signal		
							F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A070910	Undersensing		
									A0712	Pacing Problem		
									A1301	Failure to Read Input Signal		

					CT449	Generators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070910	Undersensing		
									A0712	Pacing Problem		
									A1301	Failure to Read Input Signal		
							F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A070910	Undersensing		
									A0712	Pacing Problem		
									A1301	Failure to Read Input Signal		
					CT981	Single use	E2403	No Clinical Signs, Symptoms or Conditions	A070910	Undersensing		
									A0712	Pacing Problem		
									A1301	Failure to Read Input Signal		
							F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A070910	Undersensing		
									A0712	Pacing Problem		
									A1301	Failure to Read Input Signal		
					CT983	Surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A070910	Undersensing		
									A0712	Pacing Problem		
									A1301	Failure to Read Input Signal		
							F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A070910	Undersensing		
									A0712	Pacing Problem		
									A1301	Failure to Read Input Signal		
13515	2023	All other reportable incidents	47270	Cardiac resynchronization therapy implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A0712	Pacing Problem		
									A1301	Failure to Read Input Signal		
							F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A0712	Pacing Problem		
									A1301	Failure to Read Input Signal		
13516	2023	All other reportable incidents	47267	Single-chamber implantable pacemaker, rate-responsive	CT141	Electrodes/Leads and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A0203	Defective Device		
							F26	No Health Consequences or Impact	A0203	Defective Device		
					CT145	Cardiovascular devices	E2403	No Clinical Signs, Symptoms or Conditions	A0203	Defective Device		
							F26	No Health Consequences or Impact	A0203	Defective Device		
					CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A0203	Defective Device		
							F26	No Health Consequences or Impact	A0203	Defective Device		
					CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A0203	Defective Device		
							F26	No Health Consequences or Impact	A0203	Defective Device		
					CT233	Surgical	E2403	No Clinical Signs, Symptoms or Conditions	A0203	Defective Device		
							F26	No Health Consequences or Impact	A0203	Defective Device		
					CT2406	Implantable	E2403	No Clinical Signs, Symptoms or Conditions	A0203	Defective Device		
							F26	No Health Consequences or Impact	A0203	Defective Device		
					CT2407	Active implantable	E2403	No Clinical Signs, Symptoms or Conditions	A0203	Defective Device		
							F26	No Health Consequences or Impact	A0203	Defective Device		
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A0203	Defective Device		
							F26	No Health Consequences or Impact	A0203	Defective Device		
					CT2727	Active implant devices/ancillaries	E2403	No Clinical Signs, Symptoms or Conditions	A0203	Defective Device		
							F26	No Health Consequences or Impact	A0203	Defective Device		
					CT2750	Implantable cardiac electrical stimulators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A0203	Defective Device		
							F26	No Health Consequences or Impact	A0203	Defective Device		
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A0203	Defective Device		
							F26	No Health Consequences or Impact	A0203	Defective Device		
					CT2985	Active	E2403	No Clinical Signs, Symptoms or Conditions	A0203	Defective Device		
							F26	No Health Consequences or Impact	A0203	Defective Device		
					CT3001	Implantable metal	E2403	No Clinical Signs, Symptoms or Conditions	A0203	Defective Device		
							F26	No Health Consequences or Impact	A0203	Defective Device		
					CT3002	Implantable synthetic polymer	E2403	No Clinical Signs, Symptoms or Conditions	A0203	Defective Device		
							F26	No Health Consequences or Impact	A0203	Defective Device		
					CT3006	Implantable inorganic material	E2403	No Clinical Signs, Symptoms or Conditions	A0203	Defective Device		
							F26	No Health Consequences or Impact	A0203	Defective Device		
					CT314	Electrically powered	E2403	No Clinical Signs, Symptoms or Conditions	A0203	Defective Device		
							F26	No Health Consequences or Impact	A0203	Defective Device		
					CT321	Long-term surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A0203	Defective Device		
							F26	No Health Consequences or Impact	A0203	Defective Device		
					CT449	Generators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A0203	Defective Device		
							F26	No Health Consequences or Impact	A0203	Defective Device		
					CT981	Single use	E2403	No Clinical Signs, Symptoms or Conditions	A0203	Defective Device		
							F26	No Health Consequences or Impact	A0203	Defective Device		
					CT983	Surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A0203	Defective Device		
							F26	No Health Consequences or Impact	A0203	Defective Device		
13517	2023	All other reportable incidents	37265	Dual-chamber implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E060109	Tachycardia	A0709	Device Sensing Problem		
									A071301	Failure to Deliver Shock/Stimulation		
							F05	Delay to Treatment/ Therapy	A0709	Device Sensing Problem		
									A071301	Failure to Deliver Shock/Stimulation		
							F1203	Life Threatening Illness or Injury	A0709	Device Sensing Problem		
									A071301	Failure to Deliver Shock/Stimulation		
							F23	Unexpected Medical Intervention	A0709	Device Sensing Problem		
									A071301	Failure to Deliver Shock/Stimulation		
13518	2023	All other reportable incidents	17882	Physiologic-monitoring defibrillation system	CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A0708	Power Problem		
							F27	Problem Identified During Non-Clinical Procedure	A0708	Power Problem		
13519	2023	All other reportable incidents	17882	Physiologic-monitoring defibrillation system	CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A090103	No Audible Prompt/Feedback		
							F27	Problem Identified During Non-Clinical Procedure	A090103	No Audible Prompt/Feedback		
13520	2023	All other reportable incidents	35223	Endocardial/interventricular septal pacing lead	CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A0401	Break		
									A070909	Oversensing		
									A090801	Signal Artifact/Noise		
							F26	No Health Consequences or Impact	A0401	Break		
									A070909	Oversensing		
									A090801	Signal Artifact/Noise		
13521	2023	All other reportable incidents	35223	Endocardial/interventricular septal pacing lead	CT1684	Cardiac pacemakers and associated devices	E011903	Syncope/Fainting	A051201	Device Dislodged or Dislocated	04/07/2023	04/07/2023
									A070101	Failure to Capture	04/07/2023	04/07/2023
									A070104	Unstable Capture Threshold	04/07/2023	04/07/2023
							E060106	Heart Block	A051201	Device Dislodged or Dislocated	04/07/2023	04/07/2023
									A070101	Failure to Capture	04/07/2023	04/07/2023
									A070104	Unstable Capture Threshold	04/07/2023	04/07/2023
							E060109	Tachycardia	A051201	Device Dislodged or Dislocated	04/07/2023	04/07/2023
									A070101	Failure to Capture	04/07/2023	04/07/2023

									A070104	Unstable Capture Threshold	04/07/2023	04/07/2023
							E1610	Head Injury	A051201	Device Dislodged or Dislocated	04/07/2023	04/07/2023
									A070101	Failure to Capture	04/07/2023	04/07/2023
									A070104	Unstable Capture Threshold	04/07/2023	04/07/2023
							E2007	Fall	A051201	Device Dislodged or Dislocated	04/07/2023	04/07/2023
									A070101	Failure to Capture	04/07/2023	04/07/2023
									A070104	Unstable Capture Threshold	04/07/2023	04/07/2023
							F08	Hospitalization or Prolonged Hospitalization	A051201	Device Dislodged or Dislocated	04/07/2023	04/07/2023
									A070101	Failure to Capture	04/07/2023	04/07/2023
									A070104	Unstable Capture Threshold	04/07/2023	04/07/2023
							F1905	Device Revision or Replacement	A051201	Device Dislodged or Dislocated	04/07/2023	04/07/2023
									A070101	Failure to Capture	04/07/2023	04/07/2023
									A070104	Unstable Capture Threshold	04/07/2023	04/07/2023
							F23	Unexpected Medical Intervention	A051201	Device Dislodged or Dislocated	04/07/2023	04/07/2023
									A070101	Failure to Capture	04/07/2023	04/07/2023
									A070104	Unstable Capture Threshold	04/07/2023	04/07/2023
13522	2023	All other reportable incidents	35223	Endocardial/interventricular septal pacing lead	CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A072202	Low Impedance		
							F12	Serious Injury/ Illness/ Impairment	A072202	Low Impedance		
							F23	Unexpected Medical Intervention	A072202	Low Impedance		
13523	2023	All other reportable incidents	47267	Single-chamber implantable pacemaker, rate-responsive	CT141	Electrodes/Leads and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
					CT145	Cardiovascular devices	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
					CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
					CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
					CT233	Surgical	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
					CT2406	Implantable	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
					CT2407	Active implantable	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
					CT2727	Active implant devices/ancillaries	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
					CT2750	Implantable cardiac electrical stimulators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
					CT2985	Active	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
					CT3001	Implantable metal	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
					CT3002	Implantable synthetic polymer	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
					CT3006	Implantable inorganic material	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
					CT314	Electrically powered	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
					CT321	Long-term surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
					CT449	Generators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem		

									A070504	Premature Discharge of Battery		
						F26	No Health Consequences or Impact		A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
					CT981	Single use	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
					CT983	Surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
13524	2023	Death	36241	Pacemaker/defibrillator lead	CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A0203	Defective Device		
									A072201	High Impedance		
							F02	Death	A0203	Defective Device		
									A072201	High Impedance		
13525	2023	All other reportable incidents	47265	Dual-chamber implantable pacemaker, rate-responsive	CT141	Electrodes/Leads and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A0908	Incorrect, Inadequate or Imprecise Result or Readings		
							F26	No Health Consequences or Impact	A0908	Incorrect, Inadequate or Imprecise Result or Readings		
					CT145	Cardiovascular devices	E2403	No Clinical Signs, Symptoms or Conditions	A0908	Incorrect, Inadequate or Imprecise Result or Readings		
							F26	No Health Consequences or Impact	A0908	Incorrect, Inadequate or Imprecise Result or Readings		
					CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A0908	Incorrect, Inadequate or Imprecise Result or Readings		
							F26	No Health Consequences or Impact	A0908	Incorrect, Inadequate or Imprecise Result or Readings		
					CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A0908	Incorrect, Inadequate or Imprecise Result or Readings		
							F26	No Health Consequences or Impact	A0908	Incorrect, Inadequate or Imprecise Result or Readings		
					CT233	Surgical	E2403	No Clinical Signs, Symptoms or Conditions	A0908	Incorrect, Inadequate or Imprecise Result or Readings		
							F26	No Health Consequences or Impact	A0908	Incorrect, Inadequate or Imprecise Result or Readings		
					CT2406	Implantable	E2403	No Clinical Signs, Symptoms or Conditions	A0908	Incorrect, Inadequate or Imprecise Result or Readings		
							F26	No Health Consequences or Impact	A0908	Incorrect, Inadequate or Imprecise Result or Readings		
					CT2407	Active implantable	E2403	No Clinical Signs, Symptoms or Conditions	A0908	Incorrect, Inadequate or Imprecise Result or Readings		
							F26	No Health Consequences or Impact	A0908	Incorrect, Inadequate or Imprecise Result or Readings		
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A0908	Incorrect, Inadequate or Imprecise Result or Readings		
							F26	No Health Consequences or Impact	A0908	Incorrect, Inadequate or Imprecise Result or Readings		
					CT2727	Active implant devices/ancillaries	E2403	No Clinical Signs, Symptoms or Conditions	A0908	Incorrect, Inadequate or Imprecise Result or Readings		
							F26	No Health Consequences or Impact	A0908	Incorrect, Inadequate or Imprecise Result or Readings		
					CT2750	Implantable cardiac electrical stimulators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A0908	Incorrect, Inadequate or Imprecise Result or Readings		
							F26	No Health Consequences or Impact	A0908	Incorrect, Inadequate or Imprecise Result or Readings		
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A0908	Incorrect, Inadequate or Imprecise Result or Readings		
							F26	No Health Consequences or Impact	A0908	Incorrect, Inadequate or Imprecise Result or Readings		
					CT2985	Active	E2403	No Clinical Signs, Symptoms or Conditions	A0908	Incorrect, Inadequate or Imprecise Result or Readings		
							F26	No Health Consequences or Impact	A0908	Incorrect, Inadequate or Imprecise Result or Readings		
					CT3001	Implantable metal	E2403	No Clinical Signs, Symptoms or Conditions	A0908	Incorrect, Inadequate or Imprecise Result or Readings		
							F26	No Health Consequences or Impact	A0908	Incorrect, Inadequate or Imprecise Result or Readings		
					CT3002	Implantable synthetic polymer	E2403	No Clinical Signs, Symptoms or Conditions	A0908	Incorrect, Inadequate or Imprecise Result or Readings		
							F26	No Health Consequences or Impact	A0908	Incorrect, Inadequate or Imprecise Result or Readings		
					CT3006	Implantable inorganic material	E2403	No Clinical Signs, Symptoms or Conditions	A0908	Incorrect, Inadequate or Imprecise Result or Readings		
							F26	No Health Consequences or Impact	A0908	Incorrect, Inadequate or Imprecise Result or Readings		
					CT314	Electrically powered	E2403	No Clinical Signs, Symptoms or Conditions	A0908	Incorrect, Inadequate or Imprecise Result or Readings		
							F26	No Health Consequences or Impact	A0908	Incorrect, Inadequate or Imprecise Result or Readings		
					CT321	Long-term surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A0908	Incorrect, Inadequate or Imprecise Result or Readings		
							F26	No Health Consequences or Impact	A0908	Incorrect, Inadequate or Imprecise Result or Readings		
					CT449	Generators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A0908	Incorrect, Inadequate or Imprecise Result or Readings		
							F26	No Health Consequences or Impact	A0908	Incorrect, Inadequate or Imprecise Result or Readings		
					CT981	Single use	E2403	No Clinical Signs, Symptoms or Conditions	A0908	Incorrect, Inadequate or Imprecise Result or Readings		
							F26	No Health Consequences or Impact	A0908	Incorrect, Inadequate or Imprecise Result or Readings		
					CT983	Surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A0908	Incorrect, Inadequate or Imprecise Result or Readings		
							F26	No Health Consequences or Impact	A0908	Incorrect, Inadequate or Imprecise Result or Readings		
13526	2023	All other reportable incidents	47265	Dual-chamber implantable pacemaker, rate-responsive	CT141	Electrodes/Leads and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	21/03/2023	21/03/2023
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	21/03/2023	21/03/2023

							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	21/03/2023	21/03/2023
					CT145	Cardiovascular devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	21/03/2023	21/03/2023
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	21/03/2023	21/03/2023
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	21/03/2023	21/03/2023
					CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	21/03/2023	21/03/2023
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	21/03/2023	21/03/2023
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	21/03/2023	21/03/2023
					CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	21/03/2023	21/03/2023
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	21/03/2023	21/03/2023
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	21/03/2023	21/03/2023
					CT233	Surgical	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	21/03/2023	21/03/2023
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	21/03/2023	21/03/2023
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	21/03/2023	21/03/2023
					CT2406	Implantable	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	21/03/2023	21/03/2023
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	21/03/2023	21/03/2023
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	21/03/2023	21/03/2023
					CT2407	Active implantable	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	21/03/2023	21/03/2023
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	21/03/2023	21/03/2023
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	21/03/2023	21/03/2023
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	21/03/2023	21/03/2023
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	21/03/2023	21/03/2023
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	21/03/2023	21/03/2023
					CT2727	Active implant devices/ancillaries	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	21/03/2023	21/03/2023
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	21/03/2023	21/03/2023
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	21/03/2023	21/03/2023
					CT2750	Implantable cardiac electrical stimulators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	21/03/2023	21/03/2023
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	21/03/2023	21/03/2023
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	21/03/2023	21/03/2023
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	21/03/2023	21/03/2023
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	21/03/2023	21/03/2023
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	21/03/2023	21/03/2023
					CT2985	Active	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	21/03/2023	21/03/2023
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	21/03/2023	21/03/2023
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	21/03/2023	21/03/2023
					CT3001	Implantable metal	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	21/03/2023	21/03/2023
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	21/03/2023	21/03/2023
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	21/03/2023	21/03/2023
					CT3002	Implantable synthetic polymer	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	21/03/2023	21/03/2023
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	21/03/2023	21/03/2023
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	21/03/2023	21/03/2023
					CT3006	Implantable inorganic material	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	21/03/2023	21/03/2023
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	21/03/2023	21/03/2023
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	21/03/2023	21/03/2023
					CT314	Electrically powered	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	21/03/2023	21/03/2023
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	21/03/2023	21/03/2023
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	21/03/2023	21/03/2023
					CT321	Long-term surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	21/03/2023	21/03/2023
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	21/03/2023	21/03/2023
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	21/03/2023	21/03/2023
					CT449	Generators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	21/03/2023	21/03/2023
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	21/03/2023	21/03/2023
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	21/03/2023	21/03/2023
					CT981	Single use	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	21/03/2023	21/03/2023
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	21/03/2023	21/03/2023
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	21/03/2023	21/03/2023
					CT983	Surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	21/03/2023	21/03/2023
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	21/03/2023	21/03/2023
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	21/03/2023	21/03/2023
13527	2023	All other reportable incidents	35852	Single-chamber implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
13528	2023	All other reportable incidents	47265	Dual-chamber implantable pacemaker, rate-responsive	CT141	Electrodes/Leads and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A160601	Premature Elective Replacement Indicator	26/10/2022	26/10/2022
							F1903	Device Explantation	A160601	Premature Elective Replacement Indicator	26/10/2022	26/10/2022
					CT145	Cardiovascular devices	E2403	No Clinical Signs, Symptoms or Conditions	A160601	Premature Elective Replacement Indicator	26/10/2022	26/10/2022
							F1903	Device Explantation	A160601	Premature Elective Replacement Indicator	26/10/2022	26/10/2022
					CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A160601	Premature Elective Replacement Indicator	26/10/2022	26/10/2022
							F1903	Device Explantation	A160601	Premature Elective Replacement Indicator	26/10/2022	26/10/2022
					CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A160601	Premature Elective Replacement Indicator	26/10/2022	26/10/2022
							F1903	Device Explantation	A160601	Premature Elective Replacement Indicator	26/10/2022	26/10/2022
					CT233	Surgical	E2403	No Clinical Signs, Symptoms or Conditions	A160601	Premature Elective Replacement Indicator	26/10/2022	26/10/2022
							F1903	Device Explantation	A160601	Premature Elective Replacement Indicator	26/10/2022	26/10/2022
					CT2406	Implantable	E2403	No Clinical Signs, Symptoms or Conditions	A160601	Premature Elective Replacement Indicator	26/10/2022	26/10/2022
							F1903	Device Explantation	A160601	Premature Elective Replacement Indicator	26/10/2022	26/10/2022
					CT2407	Active implantable	E2403	No Clinical Signs, Symptoms or Conditions	A160601	Premature Elective Replacement Indicator	26/10/2022	26/10/2022
							F1903	Device Explantation	A160601	Premature Elective Replacement Indicator	26/10/2022	26/10/2022
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A160601	Premature Elective Replacement Indicator	26/10/2022	26/10/2022
							F1903	Device Explantation	A160601	Premature Elective Replacement Indicator	26/10/2022	26/10/2022
					CT2727	Active implant devices/ancillaries	E2403	No Clinical Signs, Symptoms or Conditions	A160601	Premature Elective Replacement Indicator	26/10/2022	26/10/2022
							F1903	Device Explantation	A160601	Premature Elective Replacement Indicator	26/10/2022	26/10/2022
					CT2750	Implantable cardiac electrical stimulators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A160601	Premature Elective Replacement Indicator	26/10/2022	26/10/2022
							F1903	Device Explantation	A160601	Premature Elective Replacement Indicator	26/10/2022	26/10/2022
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A160601	Premature Elective Replacement Indicator	26/10/2022	26/10/2022
							F1903	Device Explantation	A160601	Premature Elective Replacement Indicator	26/10/2022	26/10/2022
					CT2985	Active	E2403	No Clinical Signs, Symptoms or Conditions	A160601	Premature Elective Replacement Indicator	26/10/2022	26/10/2022
							F1903	Device Explantation	A160601	Premature Elective Replacement Indicator	26/10/2022	26/10/2022
					CT3001	Implantable metal	E2403	No Clinical Signs, Symptoms or Conditions	A160601	Premature Elective Replacement Indicator	26/10/2022	26/10/2022
							F1903	Device Explantation	A160601	Premature Elective Replacement Indicator	26/10/2022	26/10/2022
					CT3002	Implantable synthetic polymer	E2403	No Clinical Signs, Symptoms or Conditions	A160601	Premature Elective Replacement Indicator	26/10/2022	26/10/2022
							F1903	Device Explantation	A160601	Premature Elective Replacement Indicator	26/10/2022	26/10/2022

					CT3006	Implantable inorganic material	E2403	No Clinical Signs, Symptoms or Conditions	A160601	Premature Elective Replacement Indicator	26/10/2022	26/10/2022
							F1903	Device Explantation	A160601	Premature Elective Replacement Indicator	26/10/2022	26/10/2022
					CT314	Electrically powered	E2403	No Clinical Signs, Symptoms or Conditions	A160601	Premature Elective Replacement Indicator	26/10/2022	26/10/2022
							F1903	Device Explantation	A160601	Premature Elective Replacement Indicator	26/10/2022	26/10/2022
					CT321	Long-term surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A160601	Premature Elective Replacement Indicator	26/10/2022	26/10/2022
							F1903	Device Explantation	A160601	Premature Elective Replacement Indicator	26/10/2022	26/10/2022
					CT449	Generators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A160601	Premature Elective Replacement Indicator	26/10/2022	26/10/2022
							F1903	Device Explantation	A160601	Premature Elective Replacement Indicator	26/10/2022	26/10/2022
					CT981	Single use	E2403	No Clinical Signs, Symptoms or Conditions	A160601	Premature Elective Replacement Indicator	26/10/2022	26/10/2022
							F1903	Device Explantation	A160601	Premature Elective Replacement Indicator	26/10/2022	26/10/2022
					CT983	Surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A160601	Premature Elective Replacement Indicator	26/10/2022	26/10/2022
							F1903	Device Explantation	A160601	Premature Elective Replacement Indicator	26/10/2022	26/10/2022
13529	2023	All other reportable incidents	47265	Dual-chamber implantable pacemaker, rate-responsive	CT141	Electrodes/Leads and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A1205	Loose or Intermittent Connection		
							F26	No Health Consequences or Impact	A1205	Loose or Intermittent Connection		
					CT145	Cardiovascular devices	E2403	No Clinical Signs, Symptoms or Conditions	A1205	Loose or Intermittent Connection		
							F26	No Health Consequences or Impact	A1205	Loose or Intermittent Connection		
					CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A1205	Loose or Intermittent Connection		
							F26	No Health Consequences or Impact	A1205	Loose or Intermittent Connection		
					CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A1205	Loose or Intermittent Connection		
							F26	No Health Consequences or Impact	A1205	Loose or Intermittent Connection		
					CT233	Surgical	E2403	No Clinical Signs, Symptoms or Conditions	A1205	Loose or Intermittent Connection		
							F26	No Health Consequences or Impact	A1205	Loose or Intermittent Connection		
					CT2406	Implantable	E2403	No Clinical Signs, Symptoms or Conditions	A1205	Loose or Intermittent Connection		
							F26	No Health Consequences or Impact	A1205	Loose or Intermittent Connection		
					CT2407	Active implantable	E2403	No Clinical Signs, Symptoms or Conditions	A1205	Loose or Intermittent Connection		
							F26	No Health Consequences or Impact	A1205	Loose or Intermittent Connection		
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A1205	Loose or Intermittent Connection		
							F26	No Health Consequences or Impact	A1205	Loose or Intermittent Connection		
					CT2727	Active implant devices/ancillaries	E2403	No Clinical Signs, Symptoms or Conditions	A1205	Loose or Intermittent Connection		
							F26	No Health Consequences or Impact	A1205	Loose or Intermittent Connection		
					CT2750	Implantable cardiac electrical stimulators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A1205	Loose or Intermittent Connection		
							F26	No Health Consequences or Impact	A1205	Loose or Intermittent Connection		
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A1205	Loose or Intermittent Connection		
							F26	No Health Consequences or Impact	A1205	Loose or Intermittent Connection		
					CT2985	Active	E2403	No Clinical Signs, Symptoms or Conditions	A1205	Loose or Intermittent Connection		
							F26	No Health Consequences or Impact	A1205	Loose or Intermittent Connection		
					CT3001	Implantable metal	E2403	No Clinical Signs, Symptoms or Conditions	A1205	Loose or Intermittent Connection		
							F26	No Health Consequences or Impact	A1205	Loose or Intermittent Connection		
					CT3002	Implantable synthetic polymer	E2403	No Clinical Signs, Symptoms or Conditions	A1205	Loose or Intermittent Connection		
							F26	No Health Consequences or Impact	A1205	Loose or Intermittent Connection		
					CT3006	Implantable inorganic material	E2403	No Clinical Signs, Symptoms or Conditions	A1205	Loose or Intermittent Connection		
							F26	No Health Consequences or Impact	A1205	Loose or Intermittent Connection		
					CT314	Electrically powered	E2403	No Clinical Signs, Symptoms or Conditions	A1205	Loose or Intermittent Connection		
							F26	No Health Consequences or Impact	A1205	Loose or Intermittent Connection		
					CT321	Long-term surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A1205	Loose or Intermittent Connection		
							F26	No Health Consequences or Impact	A1205	Loose or Intermittent Connection		
					CT449	Generators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A1205	Loose or Intermittent Connection		
							F26	No Health Consequences or Impact	A1205	Loose or Intermittent Connection		
					CT981	Single use	E2403	No Clinical Signs, Symptoms or Conditions	A1205	Loose or Intermittent Connection		
							F26	No Health Consequences or Impact	A1205	Loose or Intermittent Connection		
					CT983	Surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A1205	Loose or Intermittent Connection		
							F26	No Health Consequences or Impact	A1205	Loose or Intermittent Connection		
13530	2023	All other reportable incidents	35223	Endocardial/interventricular septal pacing lead	CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070909	Oversensing		
									A090801	Signal Artifact/Noise		
							F26	No Health Consequences or Impact	A070909	Oversensing		
									A090801	Signal Artifact/Noise		
13531	2023	All other reportable incidents	47263	Cardiac resynchronization therapy implantable pacemaker	CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A0908	Incorrect, Inadequate or Imprecise Result or Readings	24/07/2024	24/07/2024
							F08	Hospitalization or Prolonged Hospitalization	A0908	Incorrect, Inadequate or Imprecise Result or Readings	24/07/2024	24/07/2024
							F1905	Device Revision or Replacement	A0908	Incorrect, Inadequate or Imprecise Result or Readings	24/07/2024	24/07/2024
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A0908	Incorrect, Inadequate or Imprecise Result or Readings	24/07/2024	24/07/2024
							F08	Hospitalization or Prolonged Hospitalization	A0908	Incorrect, Inadequate or Imprecise Result or Readings	24/07/2024	24/07/2024
							F1905	Device Revision or Replacement	A0908	Incorrect, Inadequate or Imprecise Result or Readings	24/07/2024	24/07/2024
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A0908	Incorrect, Inadequate or Imprecise Result or Readings	24/07/2024	24/07/2024
							F08	Hospitalization or Prolonged Hospitalization	A0908	Incorrect, Inadequate or Imprecise Result or Readings	24/07/2024	24/07/2024
							F1905	Device Revision or Replacement	A0908	Incorrect, Inadequate or Imprecise Result or Readings	24/07/2024	24/07/2024
13532	2023	All other reportable incidents	63188	Left-ventricular wireless pacing system sensor/transmitter	CT141	Electrodes/Leads and associated devices			A0708	Power Problem		
					CT145	Cardiovascular devices			A0708	Power Problem		
					CT1577	Wireless monitoring system transmitters			A0708	Power Problem		
					CT1684	Cardiac pacemakers and associated devices			A0708	Power Problem		
					CT2317	Transmitters/Receivers/Transceivers			A0708	Power Problem		
					CT233	Surgical			A0708	Power Problem		
					CT2406	Implantable			A0708	Power Problem		
					CT2407	Active implantable			A0708	Power Problem		
					CT255	Transducers and associated devices			A0708	Power Problem		
					CT2726	Physiological monitoring/measuring devices			A0708	Power Problem		
					CT2727	Active implant devices/ancillaries			A0708	Power Problem		
					CT2750	Implantable cardiac electrical stimulators and associated devices			A0708	Power Problem		
					CT2885	Physiological monitoring/measuring system transmitters/receivers/transceivers			A0708	Power Problem		
					CT2985	Active			A0708	Power Problem		

					CT3001	Implantable metal			A0708	Power Problem		
					CT3002	Implantable synthetic polymer			A0708	Power Problem		
					CT3006	Implantable inorganic material			A0708	Power Problem		
					CT314	Electrically powered			A0708	Power Problem		
					CT321	Long-term surgical invasive			A0708	Power Problem		
					CT384	Sensors and associated devices			A0708	Power Problem		
					CT449	Generators and associated devices			A0708	Power Problem		
					CT981	Single use			A0708	Power Problem		
					CT983	Surgical invasive			A0708	Power Problem		
					CT994	Sensors			A0708	Power Problem		
13533	2023		47265	Dual-chamber implantable pacemaker, rate-responsive	CT141	Electrodes/Leads and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	21/06/2023	21/06/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	21/06/2023	21/06/2023
					CT145	Cardiovascular devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	21/06/2023	21/06/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	21/06/2023	21/06/2023
					CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	21/06/2023	21/06/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	21/06/2023	21/06/2023
					CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	21/06/2023	21/06/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	21/06/2023	21/06/2023
					CT233	Surgical	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	21/06/2023	21/06/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	21/06/2023	21/06/2023
					CT2406	Implantable	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	21/06/2023	21/06/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	21/06/2023	21/06/2023
					CT2407	Active implantable	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	21/06/2023	21/06/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	21/06/2023	21/06/2023
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	21/06/2023	21/06/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	21/06/2023	21/06/2023
					CT2727	Active implant devices/ancillaries	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	21/06/2023	21/06/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	21/06/2023	21/06/2023
					CT2750	Implantable cardiac electrical stimulators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	21/06/2023	21/06/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	21/06/2023	21/06/2023
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	21/06/2023	21/06/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	21/06/2023	21/06/2023
					CT2985	Active	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	21/06/2023	21/06/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	21/06/2023	21/06/2023
					CT3001	Implantable metal	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	21/06/2023	21/06/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	21/06/2023	21/06/2023
					CT3002	Implantable synthetic polymer	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	21/06/2023	21/06/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	21/06/2023	21/06/2023
					CT3006	Implantable inorganic material	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	21/06/2023	21/06/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	21/06/2023	21/06/2023
					CT314	Electrically powered	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	21/06/2023	21/06/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	21/06/2023	21/06/2023
					CT321	Long-term surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	21/06/2023	21/06/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	21/06/2023	21/06/2023
					CT449	Generators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	21/06/2023	21/06/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	21/06/2023	21/06/2023
					CT981	Single use	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	21/06/2023	21/06/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	21/06/2023	21/06/2023
					CT983	Surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	21/06/2023	21/06/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	21/06/2023	21/06/2023
13534	2023	All other reportable incidents	47267	Single-chamber implantable pacemaker, rate-responsive	CT141	Electrodes/Leads and associated devices	E060104	Bradycardia	A070103	Intermittent Capture	23/02/2023	23/02/2023
							F1903	Device Explantation	A070103	Intermittent Capture	23/02/2023	23/02/2023
					CT145	Cardiovascular devices	E060104	Bradycardia	A070103	Intermittent Capture	23/02/2023	23/02/2023
							F1903	Device Explantation	A070103	Intermittent Capture	23/02/2023	23/02/2023
					CT1684	Cardiac pacemakers and associated devices	E060104	Bradycardia	A070103	Intermittent Capture	23/02/2023	23/02/2023
							F1903	Device Explantation	A070103	Intermittent Capture	23/02/2023	23/02/2023
					CT1685	Cardiac pacemakers	E060104	Bradycardia	A070103	Intermittent Capture	23/02/2023	23/02/2023
							F1903	Device Explantation	A070103	Intermittent Capture	23/02/2023	23/02/2023
					CT233	Surgical	E060104	Bradycardia	A070103	Intermittent Capture	23/02/2023	23/02/2023
							F1903	Device Explantation	A070103	Intermittent Capture	23/02/2023	23/02/2023
					CT2406	Implantable	E060104	Bradycardia	A070103	Intermittent Capture	23/02/2023	23/02/2023
							F1903	Device Explantation	A070103	Intermittent Capture	23/02/2023	23/02/2023
					CT2407	Active implantable	E060104	Bradycardia	A070103	Intermittent Capture	23/02/2023	23/02/2023
							F1903	Device Explantation	A070103	Intermittent Capture	23/02/2023	23/02/2023
					CT257	Implantable cardiac pacemakers	E060104	Bradycardia	A070103	Intermittent Capture	23/02/2023	23/02/2023
							F1903	Device Explantation	A070103	Intermittent Capture	23/02/2023	23/02/2023
					CT2727	Active implant devices/ancillaries	E060104	Bradycardia	A070103	Intermittent Capture	23/02/2023	23/02/2023
							F1903	Device Explantation	A070103	Intermittent Capture	23/02/2023	23/02/2023
					CT2750	Implantable cardiac electrical stimulators and associated devices	E060104	Bradycardia	A070103	Intermittent Capture	23/02/2023	23/02/2023
							F1903	Device Explantation	A070103	Intermittent Capture	23/02/2023	23/02/2023
					CT2866	Implantable cardiac electrical stimulators	E060104	Bradycardia	A070103	Intermittent Capture	23/02/2023	23/02/2023
							F1903	Device Explantation	A070103	Intermittent Capture	23/02/2023	23/02/2023
					CT2985	Active	E060104	Bradycardia	A070103	Intermittent Capture	23/02/2023	23/02/2023
							F1903	Device Explantation	A070103	Intermittent Capture	23/02/2023	23/02/2023
					CT3001	Implantable metal	E060104	Bradycardia	A070103	Intermittent Capture	23/02/2023	23/02/2023
							F1903	Device Explantation	A070103	Intermittent Capture	23/02/2023	23/02/2023
					CT3002	Implantable synthetic polymer	E060104	Bradycardia	A070103	Intermittent Capture	23/02/2023	23/02/2023
							F1903	Device Explantation	A070103	Intermittent Capture	23/02/2023	23/02/2023
					CT3006	Implantable inorganic material	E060104	Bradycardia	A070103	Intermittent Capture	23/02/2023	23/02/2023
							F1903	Device Explantation	A070103	Intermittent Capture	23/02/2023	23/02/2023
					CT314	Electrically powered	E060104	Bradycardia	A070103	Intermittent Capture	23/02/2023	23/02/2023
							F1903	Device Explantation	A070103	Intermittent Capture	23/02/2023	23/02/2023
					CT321	Long-term surgical invasive	E060104	Bradycardia	A070103	Intermittent Capture	23/02/2023	23/02/2023
							F1903	Device Explantation	A070103	Intermittent Capture	23/02/2023	23/02/2023
					CT449	Generators and associated devices	E060104	Bradycardia	A070103	Intermittent Capture	23/02/2023	23/02/2023
							F1903	Device Explantation	A070103	Intermittent Capture	23/02/2023	23/02/2023
					CT981	Single use	E060104	Bradycardia	A070103	Intermittent Capture	23/02/2023	23/02/2023
							F1903	Device Explantation	A070103	Intermittent Capture	23/02/2023	23/02/2023

					CT983	Surgical invasive	E060104	Bradycardia	A070103	Intermittent Capture	23/02/2023	23/02/2023
							F1903	Device Explantation	A070103	Intermittent Capture	23/02/2023	23/02/2023
13535	2023	All other reportable incidents	35223	Endocardial/interventricular septal pacing lead	CT1684	Cardiac pacemakers and associated devices	E060104	Bradycardia	A070101	Failure to Capture	23/02/2023	23/02/2023
							F1903	Device Explantation	A070101	Failure to Capture	23/02/2023	23/02/2023
13536	2023	All other reportable incidents	47265	Dual-chamber implantable pacemaker, rate-responsive	CT141	Electrodes/Leads and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070910	Undersensing		
									A0712	Pacing Problem		
									A072201	High Impedance		
									A1301	Failure to Read Input Signal		
							F2203	Imaging Required	A070910	Undersensing		
									A0712	Pacing Problem		
									A072201	High Impedance		
									A1301	Failure to Read Input Signal		
							F26	No Health Consequences or Impact	A070910	Undersensing		
									A0712	Pacing Problem		
									A072201	High Impedance		
									A1301	Failure to Read Input Signal		
					CT145	Cardiovascular devices	E2403	No Clinical Signs, Symptoms or Conditions	A070910	Undersensing		
									A0712	Pacing Problem		
									A072201	High Impedance		
									A1301	Failure to Read Input Signal		
							F2203	Imaging Required	A070910	Undersensing		
									A0712	Pacing Problem		
									A072201	High Impedance		
									A1301	Failure to Read Input Signal		
							F26	No Health Consequences or Impact	A070910	Undersensing		
									A0712	Pacing Problem		
									A072201	High Impedance		
									A1301	Failure to Read Input Signal		
					CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070910	Undersensing		
									A0712	Pacing Problem		
									A072201	High Impedance		
									A1301	Failure to Read Input Signal		
							F2203	Imaging Required	A070910	Undersensing		
									A0712	Pacing Problem		
									A072201	High Impedance		
									A1301	Failure to Read Input Signal		
							F26	No Health Consequences or Impact	A070910	Undersensing		
									A0712	Pacing Problem		
									A072201	High Impedance		
									A1301	Failure to Read Input Signal		
					CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A070910	Undersensing		
									A0712	Pacing Problem		
									A072201	High Impedance		
									A1301	Failure to Read Input Signal		
							F2203	Imaging Required	A070910	Undersensing		
									A0712	Pacing Problem		
									A072201	High Impedance		
									A1301	Failure to Read Input Signal		
							F26	No Health Consequences or Impact	A070910	Undersensing		
									A0712	Pacing Problem		
									A072201	High Impedance		
									A1301	Failure to Read Input Signal		
					CT233	Surgical	E2403	No Clinical Signs, Symptoms or Conditions	A070910	Undersensing		
									A0712	Pacing Problem		
									A072201	High Impedance		
									A1301	Failure to Read Input Signal		
							F2203	Imaging Required	A070910	Undersensing		
									A0712	Pacing Problem		
									A072201	High Impedance		
									A1301	Failure to Read Input Signal		
							F26	No Health Consequences or Impact	A070910	Undersensing		
									A0712	Pacing Problem		
									A072201	High Impedance		
									A1301	Failure to Read Input Signal		
					CT2406	Implantable	E2403	No Clinical Signs, Symptoms or Conditions	A070910	Undersensing		
									A0712	Pacing Problem		
									A072201	High Impedance		
									A1301	Failure to Read Input Signal		
							F2203	Imaging Required	A070910	Undersensing		
									A0712	Pacing Problem		
									A072201	High Impedance		
									A1301	Failure to Read Input Signal		
							F26	No Health Consequences or Impact	A070910	Undersensing		
									A0712	Pacing Problem		
									A072201	High Impedance		
									A1301	Failure to Read Input Signal		
					CT2407	Active implantable	E2403	No Clinical Signs, Symptoms or Conditions	A070910	Undersensing		
									A0712	Pacing Problem		
									A072201	High Impedance		
									A1301	Failure to Read Input Signal		
							F2203	Imaging Required	A070910	Undersensing		
									A0712	Pacing Problem		
									A072201	High Impedance		
									A1301	Failure to Read Input Signal		
							F26	No Health Consequences or Impact	A070910	Undersensing		
									A0712	Pacing Problem		
									A072201	High Impedance		
									A1301	Failure to Read Input Signal		

					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A070910	Undersensing		
									A0712	Pacing Problem		
									A072201	High Impedance		
									A1301	Failure to Read Input Signal		
							F2203	Imaging Required	A070910	Undersensing		
									A0712	Pacing Problem		
									A072201	High Impedance		
									A1301	Failure to Read Input Signal		
							F26	No Health Consequences or Impact	A070910	Undersensing		
									A0712	Pacing Problem		
									A072201	High Impedance		
									A1301	Failure to Read Input Signal		
					CT2727	Active implant devices/ancillaries	E2403	No Clinical Signs, Symptoms or Conditions	A070910	Undersensing		
									A0712	Pacing Problem		
									A072201	High Impedance		
									A1301	Failure to Read Input Signal		
							F2203	Imaging Required	A070910	Undersensing		
									A0712	Pacing Problem		
									A072201	High Impedance		
									A1301	Failure to Read Input Signal		
							F26	No Health Consequences or Impact	A070910	Undersensing		
									A0712	Pacing Problem		
									A072201	High Impedance		
									A1301	Failure to Read Input Signal		
					CT2750	Implantable cardiac electrical stimulators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070910	Undersensing		
									A0712	Pacing Problem		
									A072201	High Impedance		
									A1301	Failure to Read Input Signal		
							F2203	Imaging Required	A070910	Undersensing		
									A0712	Pacing Problem		
									A072201	High Impedance		
									A1301	Failure to Read Input Signal		
							F26	No Health Consequences or Impact	A070910	Undersensing		
									A0712	Pacing Problem		
									A072201	High Impedance		
									A1301	Failure to Read Input Signal		
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A070910	Undersensing		
									A0712	Pacing Problem		
									A072201	High Impedance		
									A1301	Failure to Read Input Signal		
							F2203	Imaging Required	A070910	Undersensing		
									A0712	Pacing Problem		
									A072201	High Impedance		
									A1301	Failure to Read Input Signal		
							F26	No Health Consequences or Impact	A070910	Undersensing		
									A0712	Pacing Problem		
									A072201	High Impedance		
									A1301	Failure to Read Input Signal		
					CT2985	Active	E2403	No Clinical Signs, Symptoms or Conditions	A070910	Undersensing		
									A0712	Pacing Problem		
									A072201	High Impedance		
									A1301	Failure to Read Input Signal		
							F2203	Imaging Required	A070910	Undersensing		
									A0712	Pacing Problem		
									A072201	High Impedance		
									A1301	Failure to Read Input Signal		
							F26	No Health Consequences or Impact	A070910	Undersensing		
									A0712	Pacing Problem		
									A072201	High Impedance		
									A1301	Failure to Read Input Signal		
					CT3001	Implantable metal	E2403	No Clinical Signs, Symptoms or Conditions	A070910	Undersensing		
									A0712	Pacing Problem		
									A072201	High Impedance		
									A1301	Failure to Read Input Signal		
							F2203	Imaging Required	A070910	Undersensing		
									A0712	Pacing Problem		
									A072201	High Impedance		
									A1301	Failure to Read Input Signal		
							F26	No Health Consequences or Impact	A070910	Undersensing		
									A0712	Pacing Problem		
									A072201	High Impedance		
									A1301	Failure to Read Input Signal		
					CT3002	Implantable synthetic polymer	E2403	No Clinical Signs, Symptoms or Conditions	A070910	Undersensing		
									A0712	Pacing Problem		
									A072201	High Impedance		
									A1301	Failure to Read Input Signal		
							F2203	Imaging Required	A070910	Undersensing		
									A0712	Pacing Problem		
									A072201	High Impedance		
									A1301	Failure to Read Input Signal		
							F26	No Health Consequences or Impact	A070910	Undersensing		
									A0712	Pacing Problem		
									A072201	High Impedance		
									A1301	Failure to Read Input Signal		
					CT3006	Implantable inorganic material	E2403	No Clinical Signs, Symptoms or Conditions	A070910	Undersensing		
									A0712	Pacing Problem		
									A072201	High Impedance		
									A1301	Failure to Read Input Signal		

							F2203	Imaging Required	A070910	Undersensing		
									A0712	Pacing Problem		
									A072201	High Impedance		
									A1301	Failure to Read Input Signal		
							F26	No Health Consequences or Impact	A070910	Undersensing		
									A0712	Pacing Problem		
									A072201	High Impedance		
									A1301	Failure to Read Input Signal		
					CT314	Electrically powered	E2403	No Clinical Signs, Symptoms or Conditions	A070910	Undersensing		
									A0712	Pacing Problem		
									A072201	High Impedance		
									A1301	Failure to Read Input Signal		
							F2203	Imaging Required	A070910	Undersensing		
									A0712	Pacing Problem		
									A072201	High Impedance		
									A1301	Failure to Read Input Signal		
							F26	No Health Consequences or Impact	A070910	Undersensing		
									A0712	Pacing Problem		
									A072201	High Impedance		
									A1301	Failure to Read Input Signal		
					CT321	Long-term surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A070910	Undersensing		
									A0712	Pacing Problem		
									A072201	High Impedance		
									A1301	Failure to Read Input Signal		
							F2203	Imaging Required	A070910	Undersensing		
									A0712	Pacing Problem		
									A072201	High Impedance		
									A1301	Failure to Read Input Signal		
							F26	No Health Consequences or Impact	A070910	Undersensing		
									A0712	Pacing Problem		
									A072201	High Impedance		
									A1301	Failure to Read Input Signal		
					CT449	Generators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070910	Undersensing		
									A0712	Pacing Problem		
									A072201	High Impedance		
									A1301	Failure to Read Input Signal		
							F2203	Imaging Required	A070910	Undersensing		
									A0712	Pacing Problem		
									A072201	High Impedance		
									A1301	Failure to Read Input Signal		
							F26	No Health Consequences or Impact	A070910	Undersensing		
									A0712	Pacing Problem		
									A072201	High Impedance		
									A1301	Failure to Read Input Signal		
					CT981	Single use	E2403	No Clinical Signs, Symptoms or Conditions	A070910	Undersensing		
									A0712	Pacing Problem		
									A072201	High Impedance		
									A1301	Failure to Read Input Signal		
							F2203	Imaging Required	A070910	Undersensing		
									A0712	Pacing Problem		
									A072201	High Impedance		
									A1301	Failure to Read Input Signal		
							F26	No Health Consequences or Impact	A070910	Undersensing		
									A0712	Pacing Problem		
									A072201	High Impedance		
									A1301	Failure to Read Input Signal		
					CT983	Surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A070910	Undersensing		
									A0712	Pacing Problem		
									A072201	High Impedance		
									A1301	Failure to Read Input Signal		
							F2203	Imaging Required	A070910	Undersensing		
									A0712	Pacing Problem		
									A072201	High Impedance		
									A1301	Failure to Read Input Signal		
							F26	No Health Consequences or Impact	A070910	Undersensing		
									A0712	Pacing Problem		
									A072201	High Impedance		
									A1301	Failure to Read Input Signal		
13537	2023	All other reportable incidents	35223	Endocardial/interventricular septal pacing lead	CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A0709	Device Sensing Problem		
									A070901	Decreased Sensitivity		
									A070910	Undersensing		
							F12	Serious Injury/ Illness/ Impairment	A0709	Device Sensing Problem		
									A070901	Decreased Sensitivity		
									A070910	Undersensing		
							F23	Unexpected Medical Intervention	A0709	Device Sensing Problem		
									A070901	Decreased Sensitivity		
									A070910	Undersensing		
13538	2023	All other reportable incidents	37265	Dual-chamber implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
13539	2023	All other reportable incidents	47265	Dual-chamber implantable pacemaker, rate-responsive	CT141	Electrodes/Leads and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070909	Oversensing		
									A072201	High Impedance		
							F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A070909	Oversensing		
									A072201	High Impedance		
					CT145	Cardiovascular devices	E2403	No Clinical Signs, Symptoms or Conditions	A070909	Oversensing		
									A072201	High Impedance		
							F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A070909	Oversensing		

									A072201	High Impedance		
					CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070909	Oversensing		
									A072201	High Impedance		
							F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A070909	Oversensing		
									A072201	High Impedance		
					CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A070909	Oversensing		
									A072201	High Impedance		
							F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A070909	Oversensing		
									A072201	High Impedance		
					CT233	Surgical	E2403	No Clinical Signs, Symptoms or Conditions	A070909	Oversensing		
									A072201	High Impedance		
							F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A070909	Oversensing		
									A072201	High Impedance		
					CT2406	Implantable	E2403	No Clinical Signs, Symptoms or Conditions	A070909	Oversensing		
									A072201	High Impedance		
							F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A070909	Oversensing		
									A072201	High Impedance		
					CT2407	Active implantable	E2403	No Clinical Signs, Symptoms or Conditions	A070909	Oversensing		
									A072201	High Impedance		
							F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A070909	Oversensing		
									A072201	High Impedance		
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A070909	Oversensing		
									A072201	High Impedance		
							F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A070909	Oversensing		
									A072201	High Impedance		
					CT2727	Active implant devices/ancillaries	E2403	No Clinical Signs, Symptoms or Conditions	A070909	Oversensing		
									A072201	High Impedance		
							F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A070909	Oversensing		
									A072201	High Impedance		
					CT2750	Implantable cardiac electrical stimulators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070909	Oversensing		
									A072201	High Impedance		
							F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A070909	Oversensing		
									A072201	High Impedance		
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A070909	Oversensing		
									A072201	High Impedance		
							F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A070909	Oversensing		
									A072201	High Impedance		
					CT2985	Active	E2403	No Clinical Signs, Symptoms or Conditions	A070909	Oversensing		
									A072201	High Impedance		
							F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A070909	Oversensing		
									A072201	High Impedance		
					CT3001	Implantable metal	E2403	No Clinical Signs, Symptoms or Conditions	A070909	Oversensing		
									A072201	High Impedance		
							F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A070909	Oversensing		
									A072201	High Impedance		
					CT3002	Implantable synthetic polymer	E2403	No Clinical Signs, Symptoms or Conditions	A070909	Oversensing		
									A072201	High Impedance		
							F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A070909	Oversensing		
									A072201	High Impedance		
					CT3006	Implantable inorganic material	E2403	No Clinical Signs, Symptoms or Conditions	A070909	Oversensing		
									A072201	High Impedance		
							F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A070909	Oversensing		
									A072201	High Impedance		
					CT314	Electrically powered	E2403	No Clinical Signs, Symptoms or Conditions	A070909	Oversensing		
									A072201	High Impedance		
							F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A070909	Oversensing		
									A072201	High Impedance		
					CT321	Long-term surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A070909	Oversensing		
									A072201	High Impedance		
							F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A070909	Oversensing		
									A072201	High Impedance		
					CT449	Generators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070909	Oversensing		
									A072201	High Impedance		
							F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A070909	Oversensing		
									A072201	High Impedance		
					CT981	Single use	E2403	No Clinical Signs, Symptoms or Conditions	A070909	Oversensing		
									A072201	High Impedance		
							F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A070909	Oversensing		
									A072201	High Impedance		
					CT983	Surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A070909	Oversensing		
									A072201	High Impedance		
							F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A070909	Oversensing		
									A072201	High Impedance		
13540	2023	All other reportable incidents	47265	Dual-chamber implantable pacemaker, rate-responsive	CT141	Electrodes/Leads and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem		
							F26	No Health Consequences or Impact	A0705	Battery Problem		
					CT145	Cardiovascular devices	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem		
							F26	No Health Consequences or Impact	A0705	Battery Problem		
					CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem		
							F26	No Health Consequences or Impact	A0705	Battery Problem		
					CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem		
							F26	No Health Consequences or Impact	A0705	Battery Problem		
					CT233	Surgical	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem		
							F26	No Health Consequences or Impact	A0705	Battery Problem		
					CT2406	Implantable	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem		
							F26	No Health Consequences or Impact	A0705	Battery Problem		
					CT2407	Active implantable	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem		
							F26	No Health Consequences or Impact	A0705	Battery Problem		
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem		

						F26	No Health Consequences or Impact	A0705	Battery Problem			
					CT2727	Active implant devices/ancillaries	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem		
							F26	No Health Consequences or Impact	A0705	Battery Problem		
					CT2750	Implantable cardiac electrical stimulators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem		
							F26	No Health Consequences or Impact	A0705	Battery Problem		
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem		
							F26	No Health Consequences or Impact	A0705	Battery Problem		
					CT2985	Active	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem		
							F26	No Health Consequences or Impact	A0705	Battery Problem		
					CT3001	Implantable metal	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem		
							F26	No Health Consequences or Impact	A0705	Battery Problem		
					CT3002	Implantable synthetic polymer	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem		
							F26	No Health Consequences or Impact	A0705	Battery Problem		
					CT3006	Implantable inorganic material	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem		
							F26	No Health Consequences or Impact	A0705	Battery Problem		
					CT314	Electrically powered	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem		
							F26	No Health Consequences or Impact	A0705	Battery Problem		
					CT321	Long-term surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem		
							F26	No Health Consequences or Impact	A0705	Battery Problem		
					CT449	Generators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem		
							F26	No Health Consequences or Impact	A0705	Battery Problem		
					CT981	Single use	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem		
							F26	No Health Consequences or Impact	A0705	Battery Problem		
					CT983	Surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem		
							F26	No Health Consequences or Impact	A0705	Battery Problem		
13541	2023	All other reportable incidents	60910	Coronary venous pacing lead	CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A0722	Impedance Problem		
							F26	No Health Consequences or Impact	A0722	Impedance Problem		
13542	2023		47265	Dual-chamber implantable pacemaker, rate-responsive	CT141	Electrodes/Leads and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	08/06/2023	08/06/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	08/06/2023	08/06/2023
					CT145	Cardiovascular devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	08/06/2023	08/06/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	08/06/2023	08/06/2023
					CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	08/06/2023	08/06/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	08/06/2023	08/06/2023
					CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	08/06/2023	08/06/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	08/06/2023	08/06/2023
					CT233	Surgical	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	08/06/2023	08/06/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	08/06/2023	08/06/2023
					CT2406	Implantable	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	08/06/2023	08/06/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	08/06/2023	08/06/2023
					CT2407	Active implantable	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	08/06/2023	08/06/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	08/06/2023	08/06/2023
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	08/06/2023	08/06/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	08/06/2023	08/06/2023
					CT2727	Active implant devices/ancillaries	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	08/06/2023	08/06/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	08/06/2023	08/06/2023
					CT2750	Implantable cardiac electrical stimulators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	08/06/2023	08/06/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	08/06/2023	08/06/2023
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	08/06/2023	08/06/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	08/06/2023	08/06/2023
					CT2985	Active	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	08/06/2023	08/06/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	08/06/2023	08/06/2023
					CT3001	Implantable metal	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	08/06/2023	08/06/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	08/06/2023	08/06/2023
					CT3002	Implantable synthetic polymer	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	08/06/2023	08/06/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	08/06/2023	08/06/2023
					CT3006	Implantable inorganic material	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	08/06/2023	08/06/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	08/06/2023	08/06/2023
					CT314	Electrically powered	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	08/06/2023	08/06/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	08/06/2023	08/06/2023
					CT321	Long-term surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	08/06/2023	08/06/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	08/06/2023	08/06/2023
					CT449	Generators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	08/06/2023	08/06/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	08/06/2023	08/06/2023
					CT981	Single use	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	08/06/2023	08/06/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	08/06/2023	08/06/2023
					CT983	Surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	08/06/2023	08/06/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	08/06/2023	08/06/2023
13543	2023		47265	Dual-chamber implantable pacemaker, rate-responsive	CT141	Electrodes/Leads and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	07/06/2023	07/06/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	07/06/2023	07/06/2023
					CT145	Cardiovascular devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	07/06/2023	07/06/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	07/06/2023	07/06/2023
					CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	07/06/2023	07/06/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	07/06/2023	07/06/2023
					CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	07/06/2023	07/06/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	07/06/2023	07/06/2023
					CT233	Surgical	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	07/06/2023	07/06/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	07/06/2023	07/06/2023
					CT2406	Implantable	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	07/06/2023	07/06/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	07/06/2023	07/06/2023
					CT2407	Active implantable	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	07/06/2023	07/06/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	07/06/2023	07/06/2023
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	07/06/2023	07/06/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	07/06/2023	07/06/2023
					CT2727	Active implant devices/ancillaries	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	07/06/2023	07/06/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	07/06/2023	07/06/2023
					CT2750	Implantable cardiac electrical stimulators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	07/06/2023	07/06/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	07/06/2023	07/06/2023
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	07/06/2023	07/06/2023

							F1901	Additional Surgery	A24	No Apparent Device Problem	07/06/2023	07/06/2023
					CT2985	Active	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	07/06/2023	07/06/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	07/06/2023	07/06/2023
					CT3001	Implantable metal	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	07/06/2023	07/06/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	07/06/2023	07/06/2023
					CT3002	Implantable synthetic polymer	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	07/06/2023	07/06/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	07/06/2023	07/06/2023
					CT3006	Implantable inorganic material	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	07/06/2023	07/06/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	07/06/2023	07/06/2023
					CT314	Electrically powered	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	07/06/2023	07/06/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	07/06/2023	07/06/2023
					CT321	Long-term surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	07/06/2023	07/06/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	07/06/2023	07/06/2023
					CT449	Generators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	07/06/2023	07/06/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	07/06/2023	07/06/2023
					CT981	Single use	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	07/06/2023	07/06/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	07/06/2023	07/06/2023
					CT983	Surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	07/06/2023	07/06/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	07/06/2023	07/06/2023
13544	2023	All other reportable incidents	47265	Dual-chamber implantable pacemaker, rate-responsive	CT141	Electrodes/Leads and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem		
									A072201	High Impedance		
									A1301	Failure to Read Input Signal		
							F26	No Health Consequences or Impact	A0705	Battery Problem		
									A072201	High Impedance		
									A1301	Failure to Read Input Signal		
					CT145	Cardiovascular devices	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem		
									A072201	High Impedance		
									A1301	Failure to Read Input Signal		
							F26	No Health Consequences or Impact	A0705	Battery Problem		
									A072201	High Impedance		
									A1301	Failure to Read Input Signal		
					CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem		
									A072201	High Impedance		
									A1301	Failure to Read Input Signal		
							F26	No Health Consequences or Impact	A0705	Battery Problem		
									A072201	High Impedance		
									A1301	Failure to Read Input Signal		
					CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem		
									A072201	High Impedance		
									A1301	Failure to Read Input Signal		
							F26	No Health Consequences or Impact	A0705	Battery Problem		
									A072201	High Impedance		
									A1301	Failure to Read Input Signal		
					CT233	Surgical	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem		
									A072201	High Impedance		
									A1301	Failure to Read Input Signal		
							F26	No Health Consequences or Impact	A0705	Battery Problem		
									A072201	High Impedance		
									A1301	Failure to Read Input Signal		
					CT2406	Implantable	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem		
									A072201	High Impedance		
									A1301	Failure to Read Input Signal		
							F26	No Health Consequences or Impact	A0705	Battery Problem		
									A072201	High Impedance		
									A1301	Failure to Read Input Signal		
					CT2407	Active implantable	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem		
									A072201	High Impedance		
									A1301	Failure to Read Input Signal		
							F26	No Health Consequences or Impact	A0705	Battery Problem		
									A072201	High Impedance		
									A1301	Failure to Read Input Signal		
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem		
									A072201	High Impedance		
									A1301	Failure to Read Input Signal		
							F26	No Health Consequences or Impact	A0705	Battery Problem		
									A072201	High Impedance		
									A1301	Failure to Read Input Signal		
					CT2727	Active implant devices/ancillaries	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem		
									A072201	High Impedance		
									A1301	Failure to Read Input Signal		
							F26	No Health Consequences or Impact	A0705	Battery Problem		
									A072201	High Impedance		
									A1301	Failure to Read Input Signal		
					CT2750	Implantable cardiac electrical stimulators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem		
									A072201	High Impedance		
									A1301	Failure to Read Input Signal		
							F26	No Health Consequences or Impact	A0705	Battery Problem		
									A072201	High Impedance		
									A1301	Failure to Read Input Signal		
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem		
									A072201	High Impedance		
									A1301	Failure to Read Input Signal		
							F26	No Health Consequences or Impact	A0705	Battery Problem		
									A072201	High Impedance		
									A1301	Failure to Read Input Signal		
					CT2985	Active	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem		
									A072201	High Impedance		
									A1301	Failure to Read Input Signal		

							F26	No Health Consequences or Impact	A0705	Battery Problem		
									A072201	High Impedance		
									A1301	Failure to Read Input Signal		
					CT3001	Implantable metal	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem		
									A072201	High Impedance		
									A1301	Failure to Read Input Signal		
							F26	No Health Consequences or Impact	A0705	Battery Problem		
									A072201	High Impedance		
									A1301	Failure to Read Input Signal		
					CT3002	Implantable synthetic polymer	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem		
									A072201	High Impedance		
									A1301	Failure to Read Input Signal		
							F26	No Health Consequences or Impact	A0705	Battery Problem		
									A072201	High Impedance		
									A1301	Failure to Read Input Signal		
					CT3006	Implantable inorganic material	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem		
									A072201	High Impedance		
									A1301	Failure to Read Input Signal		
							F26	No Health Consequences or Impact	A0705	Battery Problem		
									A072201	High Impedance		
									A1301	Failure to Read Input Signal		
					CT314	Electrically powered	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem		
									A072201	High Impedance		
									A1301	Failure to Read Input Signal		
							F26	No Health Consequences or Impact	A0705	Battery Problem		
									A072201	High Impedance		
									A1301	Failure to Read Input Signal		
					CT321	Long-term surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem		
									A072201	High Impedance		
									A1301	Failure to Read Input Signal		
							F26	No Health Consequences or Impact	A0705	Battery Problem		
									A072201	High Impedance		
									A1301	Failure to Read Input Signal		
					CT449	Generators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem		
									A072201	High Impedance		
									A1301	Failure to Read Input Signal		
							F26	No Health Consequences or Impact	A0705	Battery Problem		
									A072201	High Impedance		
									A1301	Failure to Read Input Signal		
					CT981	Single use	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem		
									A072201	High Impedance		
									A1301	Failure to Read Input Signal		
							F26	No Health Consequences or Impact	A0705	Battery Problem		
									A072201	High Impedance		
									A1301	Failure to Read Input Signal		
					CT983	Surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem		
									A072201	High Impedance		
									A1301	Failure to Read Input Signal		
							F26	No Health Consequences or Impact	A0705	Battery Problem		
									A072201	High Impedance		
									A1301	Failure to Read Input Signal		
13545	2023	All other reportable incidents	47263	Cardiac resynchronization therapy implantable pacemaker	CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A0203	Defective Device	25/05/2023	25/05/2023
							F08	Hospitalization or Prolonged Hospitalization	A0203	Defective Device	25/05/2023	25/05/2023
							F1905	Device Revision or Replacement	A0203	Defective Device	25/05/2023	25/05/2023
					CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A0203	Defective Device	25/05/2023	25/05/2023
							F08	Hospitalization or Prolonged Hospitalization	A0203	Defective Device	25/05/2023	25/05/2023
							F1905	Device Revision or Replacement	A0203	Defective Device	25/05/2023	25/05/2023
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A0203	Defective Device	25/05/2023	25/05/2023
							F08	Hospitalization or Prolonged Hospitalization	A0203	Defective Device	25/05/2023	25/05/2023
							F1905	Device Revision or Replacement	A0203	Defective Device	25/05/2023	25/05/2023
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A0203	Defective Device	25/05/2023	25/05/2023
							F08	Hospitalization or Prolonged Hospitalization	A0203	Defective Device	25/05/2023	25/05/2023
							F1905	Device Revision or Replacement	A0203	Defective Device	25/05/2023	25/05/2023
13546	2023	All other reportable incidents	47265	Dual-chamber implantable pacemaker, rate-responsive	CT141	Electrodes/Leads and associated devices	E0112	Dizziness	A071102	Failure to Interrogate	28/03/2023	28/03/2023
									A071204	No Pacing	28/03/2023	28/03/2023
							E0601	Arrhythmia	A071102	Failure to Interrogate	28/03/2023	28/03/2023
									A071204	No Pacing	28/03/2023	28/03/2023
							F1901	Additional Surgery	A071102	Failure to Interrogate	28/03/2023	28/03/2023
									A071204	No Pacing	28/03/2023	28/03/2023
					CT145	Cardiovascular devices	E0112	Dizziness	A071102	Failure to Interrogate	28/03/2023	28/03/2023
									A071204	No Pacing	28/03/2023	28/03/2023
							E0601	Arrhythmia	A071102	Failure to Interrogate	28/03/2023	28/03/2023
									A071204	No Pacing	28/03/2023	28/03/2023
							F1901	Additional Surgery	A071102	Failure to Interrogate	28/03/2023	28/03/2023
									A071204	No Pacing	28/03/2023	28/03/2023
					CT1684	Cardiac pacemakers and associated devices	E0112	Dizziness	A071102	Failure to Interrogate	28/03/2023	28/03/2023
									A071204	No Pacing	28/03/2023	28/03/2023
							E0601	Arrhythmia	A071102	Failure to Interrogate	28/03/2023	28/03/2023
									A071204	No Pacing	28/03/2023	28/03/2023
							F1901	Additional Surgery	A071102	Failure to Interrogate	28/03/2023	28/03/2023
									A071204	No Pacing	28/03/2023	28/03/2023
					CT1685	Cardiac pacemakers	E0112	Dizziness	A071102	Failure to Interrogate	28/03/2023	28/03/2023
									A071204	No Pacing	28/03/2023	28/03/2023
							E0601	Arrhythmia	A071102	Failure to Interrogate	28/03/2023	28/03/2023
									A071204	No Pacing	28/03/2023	28/03/2023
							F1901	Additional Surgery	A071102	Failure to Interrogate	28/03/2023	28/03/2023
									A071204	No Pacing	28/03/2023	28/03/2023
					CT233	Surgical	E0112	Dizziness	A071102	Failure to Interrogate	28/03/2023	28/03/2023

									A071204	No Pacing	28/03/2023	28/03/2023
							E0601	Arrhythmia	A071102	Failure to Interrogate	28/03/2023	28/03/2023
									A071204	No Pacing	28/03/2023	28/03/2023
							F1901	Additional Surgery	A071102	Failure to Interrogate	28/03/2023	28/03/2023
									A071204	No Pacing	28/03/2023	28/03/2023
					CT2406	Implantable	E0112	Dizziness	A071102	Failure to Interrogate	28/03/2023	28/03/2023
									A071204	No Pacing	28/03/2023	28/03/2023
							E0601	Arrhythmia	A071102	Failure to Interrogate	28/03/2023	28/03/2023
									A071204	No Pacing	28/03/2023	28/03/2023
							F1901	Additional Surgery	A071102	Failure to Interrogate	28/03/2023	28/03/2023
									A071204	No Pacing	28/03/2023	28/03/2023
					CT2407	Active implantable	E0112	Dizziness	A071102	Failure to Interrogate	28/03/2023	28/03/2023
									A071204	No Pacing	28/03/2023	28/03/2023
							E0601	Arrhythmia	A071102	Failure to Interrogate	28/03/2023	28/03/2023
									A071204	No Pacing	28/03/2023	28/03/2023
							F1901	Additional Surgery	A071102	Failure to Interrogate	28/03/2023	28/03/2023
									A071204	No Pacing	28/03/2023	28/03/2023
					CT257	Implantable cardiac pacemakers	E0112	Dizziness	A071102	Failure to Interrogate	28/03/2023	28/03/2023
									A071204	No Pacing	28/03/2023	28/03/2023
							E0601	Arrhythmia	A071102	Failure to Interrogate	28/03/2023	28/03/2023
									A071204	No Pacing	28/03/2023	28/03/2023
							F1901	Additional Surgery	A071102	Failure to Interrogate	28/03/2023	28/03/2023
									A071204	No Pacing	28/03/2023	28/03/2023
					CT2727	Active implant devices/ancillaries	E0112	Dizziness	A071102	Failure to Interrogate	28/03/2023	28/03/2023
									A071204	No Pacing	28/03/2023	28/03/2023
							E0601	Arrhythmia	A071102	Failure to Interrogate	28/03/2023	28/03/2023
									A071204	No Pacing	28/03/2023	28/03/2023
							F1901	Additional Surgery	A071102	Failure to Interrogate	28/03/2023	28/03/2023
									A071204	No Pacing	28/03/2023	28/03/2023
					CT2750	Implantable cardiac electrical stimulators and associated devices	E0112	Dizziness	A071102	Failure to Interrogate	28/03/2023	28/03/2023
									A071204	No Pacing	28/03/2023	28/03/2023
							E0601	Arrhythmia	A071102	Failure to Interrogate	28/03/2023	28/03/2023
									A071204	No Pacing	28/03/2023	28/03/2023
							F1901	Additional Surgery	A071102	Failure to Interrogate	28/03/2023	28/03/2023
									A071204	No Pacing	28/03/2023	28/03/2023
					CT2866	Implantable cardiac electrical stimulators	E0112	Dizziness	A071102	Failure to Interrogate	28/03/2023	28/03/2023
									A071204	No Pacing	28/03/2023	28/03/2023
							E0601	Arrhythmia	A071102	Failure to Interrogate	28/03/2023	28/03/2023
									A071204	No Pacing	28/03/2023	28/03/2023
							F1901	Additional Surgery	A071102	Failure to Interrogate	28/03/2023	28/03/2023
									A071204	No Pacing	28/03/2023	28/03/2023
					CT2985	Active	E0112	Dizziness	A071102	Failure to Interrogate	28/03/2023	28/03/2023
									A071204	No Pacing	28/03/2023	28/03/2023
							E0601	Arrhythmia	A071102	Failure to Interrogate	28/03/2023	28/03/2023
									A071204	No Pacing	28/03/2023	28/03/2023
							F1901	Additional Surgery	A071102	Failure to Interrogate	28/03/2023	28/03/2023
									A071204	No Pacing	28/03/2023	28/03/2023
					CT3001	Implantable metal	E0112	Dizziness	A071102	Failure to Interrogate	28/03/2023	28/03/2023
									A071204	No Pacing	28/03/2023	28/03/2023
							E0601	Arrhythmia	A071102	Failure to Interrogate	28/03/2023	28/03/2023
									A071204	No Pacing	28/03/2023	28/03/2023
							F1901	Additional Surgery	A071102	Failure to Interrogate	28/03/2023	28/03/2023
									A071204	No Pacing	28/03/2023	28/03/2023
					CT3002	Implantable synthetic polymer	E0112	Dizziness	A071102	Failure to Interrogate	28/03/2023	28/03/2023
									A071204	No Pacing	28/03/2023	28/03/2023
							E0601	Arrhythmia	A071102	Failure to Interrogate	28/03/2023	28/03/2023
									A071204	No Pacing	28/03/2023	28/03/2023
							F1901	Additional Surgery	A071102	Failure to Interrogate	28/03/2023	28/03/2023
									A071204	No Pacing	28/03/2023	28/03/2023
					CT3006	Implantable inorganic material	E0112	Dizziness	A071102	Failure to Interrogate	28/03/2023	28/03/2023
									A071204	No Pacing	28/03/2023	28/03/2023
							E0601	Arrhythmia	A071102	Failure to Interrogate	28/03/2023	28/03/2023
									A071204	No Pacing	28/03/2023	28/03/2023
							F1901	Additional Surgery	A071102	Failure to Interrogate	28/03/2023	28/03/2023
									A071204	No Pacing	28/03/2023	28/03/2023
					CT314	Electrically powered	E0112	Dizziness	A071102	Failure to Interrogate	28/03/2023	28/03/2023
									A071204	No Pacing	28/03/2023	28/03/2023
							E0601	Arrhythmia	A071102	Failure to Interrogate	28/03/2023	28/03/2023
									A071204	No Pacing	28/03/2023	28/03/2023
							F1901	Additional Surgery	A071102	Failure to Interrogate	28/03/2023	28/03/2023
									A071204	No Pacing	28/03/2023	28/03/2023
					CT321	Long-term surgical invasive	E0112	Dizziness	A071102	Failure to Interrogate	28/03/2023	28/03/2023
									A071204	No Pacing	28/03/2023	28/03/2023
							E0601	Arrhythmia	A071102	Failure to Interrogate	28/03/2023	28/03/2023
									A071204	No Pacing	28/03/2023	28/03/2023
							F1901	Additional Surgery	A071102	Failure to Interrogate	28/03/2023	28/03/2023
									A071204	No Pacing	28/03/2023	28/03/2023
					CT449	Generators and associated devices	E0112	Dizziness	A071102	Failure to Interrogate	28/03/2023	28/03/2023
									A071204	No Pacing	28/03/2023	28/03/2023
							E0601	Arrhythmia	A071102	Failure to Interrogate	28/03/2023	28/03/2023
									A071204	No Pacing	28/03/2023	28/03/2023
							F1901	Additional Surgery	A071102	Failure to Interrogate	28/03/2023	28/03/2023
									A071204	No Pacing	28/03/2023	28/03/2023
					CT981	Single use	E0112	Dizziness	A071102	Failure to Interrogate	28/03/2023	28/03/2023
									A071204	No Pacing	28/03/2023	28/03/2023
							E0601	Arrhythmia	A071102	Failure to Interrogate	28/03/2023	28/03/2023
									A071204	No Pacing	28/03/2023	28/03/2023
							F1901	Additional Surgery	A071102	Failure to Interrogate	28/03/2023	28/03/2023

								A071204	No Pacing	28/03/2023	28/03/2023
				CT983	Surgical invasive	E0112	Dizziness	A071102	Failure to Interrogate	28/03/2023	28/03/2023
								A071204	No Pacing	28/03/2023	28/03/2023
						E0601	Arrhythmia	A071102	Failure to Interrogate	28/03/2023	28/03/2023
								A071204	No Pacing	28/03/2023	28/03/2023
						F1901	Additional Surgery	A071102	Failure to Interrogate	28/03/2023	28/03/2023
								A071204	No Pacing	28/03/2023	28/03/2023
13547	2023	All other reportable incidents	47265	Dual-chamber implantable pacemaker, rate-responsive	CT141	Electrodes/Leads and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A1606	Premature Indicator Activation		
							F1905	Device Revision or Replacement	A1606	Premature Indicator Activation		
					CT145	Cardiovascular devices	E2403	No Clinical Signs, Symptoms or Conditions	A1606	Premature Indicator Activation		
							F1905	Device Revision or Replacement	A1606	Premature Indicator Activation		
					CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A1606	Premature Indicator Activation		
							F1905	Device Revision or Replacement	A1606	Premature Indicator Activation		
					CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A1606	Premature Indicator Activation		
							F1905	Device Revision or Replacement	A1606	Premature Indicator Activation		
					CT233	Surgical	E2403	No Clinical Signs, Symptoms or Conditions	A1606	Premature Indicator Activation		
							F1905	Device Revision or Replacement	A1606	Premature Indicator Activation		
					CT2406	Implantable	E2403	No Clinical Signs, Symptoms or Conditions	A1606	Premature Indicator Activation		
							F1905	Device Revision or Replacement	A1606	Premature Indicator Activation		
					CT2407	Active implantable	E2403	No Clinical Signs, Symptoms or Conditions	A1606	Premature Indicator Activation		
							F1905	Device Revision or Replacement	A1606	Premature Indicator Activation		
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A1606	Premature Indicator Activation		
							F1905	Device Revision or Replacement	A1606	Premature Indicator Activation		
					CT2727	Active implant devices/ancillaries	E2403	No Clinical Signs, Symptoms or Conditions	A1606	Premature Indicator Activation		
							F1905	Device Revision or Replacement	A1606	Premature Indicator Activation		
					CT2750	Implantable cardiac electrical stimulators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A1606	Premature Indicator Activation		
							F1905	Device Revision or Replacement	A1606	Premature Indicator Activation		
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A1606	Premature Indicator Activation		
							F1905	Device Revision or Replacement	A1606	Premature Indicator Activation		
					CT2985	Active	E2403	No Clinical Signs, Symptoms or Conditions	A1606	Premature Indicator Activation		
							F1905	Device Revision or Replacement	A1606	Premature Indicator Activation		
					CT3001	Implantable metal	E2403	No Clinical Signs, Symptoms or Conditions	A1606	Premature Indicator Activation		
							F1905	Device Revision or Replacement	A1606	Premature Indicator Activation		
					CT3002	Implantable synthetic polymer	E2403	No Clinical Signs, Symptoms or Conditions	A1606	Premature Indicator Activation		
							F1905	Device Revision or Replacement	A1606	Premature Indicator Activation		
					CT3006	Implantable inorganic material	E2403	No Clinical Signs, Symptoms or Conditions	A1606	Premature Indicator Activation		
							F1905	Device Revision or Replacement	A1606	Premature Indicator Activation		
					CT314	Electrically powered	E2403	No Clinical Signs, Symptoms or Conditions	A1606	Premature Indicator Activation		
							F1905	Device Revision or Replacement	A1606	Premature Indicator Activation		
					CT321	Long-term surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A1606	Premature Indicator Activation		
							F1905	Device Revision or Replacement	A1606	Premature Indicator Activation		
					CT449	Generators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A1606	Premature Indicator Activation		
							F1905	Device Revision or Replacement	A1606	Premature Indicator Activation		
					CT981	Single use	E2403	No Clinical Signs, Symptoms or Conditions	A1606	Premature Indicator Activation		
							F1905	Device Revision or Replacement	A1606	Premature Indicator Activation		
					CT983	Surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A1606	Premature Indicator Activation		
							F1905	Device Revision or Replacement	A1606	Premature Indicator Activation		
13548	2023	All other reportable incidents	47265	Dual-chamber implantable pacemaker, rate-responsive	CT141	Electrodes/Leads and associated devices	E233001	Chest Pain	A05	Mechanical Problem	10/07/2023	10/07/2023
							F08	Hospitalization or Prolonged Hospitalization	A05	Mechanical Problem	10/07/2023	10/07/2023
							F12	Serious Injury/ Illness/ Impairment	A05	Mechanical Problem	10/07/2023	10/07/2023
							F1905	Device Revision or Replacement	A05	Mechanical Problem	10/07/2023	10/07/2023
					CT145	Cardiovascular devices	E233001	Chest Pain	A05	Mechanical Problem	10/07/2023	10/07/2023
							F08	Hospitalization or Prolonged Hospitalization	A05	Mechanical Problem	10/07/2023	10/07/2023
							F12	Serious Injury/ Illness/ Impairment	A05	Mechanical Problem	10/07/2023	10/07/2023
							F1905	Device Revision or Replacement	A05	Mechanical Problem	10/07/2023	10/07/2023
					CT1684	Cardiac pacemakers and associated devices	E233001	Chest Pain	A05	Mechanical Problem	10/07/2023	10/07/2023
							F08	Hospitalization or Prolonged Hospitalization	A05	Mechanical Problem	10/07/2023	10/07/2023
							F12	Serious Injury/ Illness/ Impairment	A05	Mechanical Problem	10/07/2023	10/07/2023
							F1905	Device Revision or Replacement	A05	Mechanical Problem	10/07/2023	10/07/2023
					CT1685	Cardiac pacemakers	E233001	Chest Pain	A05	Mechanical Problem	10/07/2023	10/07/2023
							F08	Hospitalization or Prolonged Hospitalization	A05	Mechanical Problem	10/07/2023	10/07/2023
							F12	Serious Injury/ Illness/ Impairment	A05	Mechanical Problem	10/07/2023	10/07/2023
							F1905	Device Revision or Replacement	A05	Mechanical Problem	10/07/2023	10/07/2023
					CT233	Surgical	E233001	Chest Pain	A05	Mechanical Problem	10/07/2023	10/07/2023
							F08	Hospitalization or Prolonged Hospitalization	A05	Mechanical Problem	10/07/2023	10/07/2023
							F12	Serious Injury/ Illness/ Impairment	A05	Mechanical Problem	10/07/2023	10/07/2023
							F1905	Device Revision or Replacement	A05	Mechanical Problem	10/07/2023	10/07/2023
					CT2406	Implantable	E233001	Chest Pain	A05	Mechanical Problem	10/07/2023	10/07/2023
							F08	Hospitalization or Prolonged Hospitalization	A05	Mechanical Problem	10/07/2023	10/07/2023
							F12	Serious Injury/ Illness/ Impairment	A05	Mechanical Problem	10/07/2023	10/07/2023
							F1905	Device Revision or Replacement	A05	Mechanical Problem	10/07/2023	10/07/2023
					CT2407	Active implantable	E233001	Chest Pain	A05	Mechanical Problem	10/07/2023	10/07/2023
							F08	Hospitalization or Prolonged Hospitalization	A05	Mechanical Problem	10/07/2023	10/07/2023
							F12	Serious Injury/ Illness/ Impairment	A05	Mechanical Problem	10/07/2023	10/07/2023
							F1905	Device Revision or Replacement	A05	Mechanical Problem	10/07/2023	10/07/2023
					CT257	Implantable cardiac pacemakers	E233001	Chest Pain	A05	Mechanical Problem	10/07/2023	10/07/2023
							F08	Hospitalization or Prolonged Hospitalization	A05	Mechanical Problem	10/07/2023	10/07/2023
							F12	Serious Injury/ Illness/ Impairment	A05	Mechanical Problem	10/07/2023	10/07/2023
							F1905	Device Revision or Replacement	A05	Mechanical Problem	10/07/2023	10/07/2023
					CT2727	Active implant devices/ancillaries	E233001	Chest Pain	A05	Mechanical Problem	10/07/2023	10/07/2023
							F08	Hospitalization or Prolonged Hospitalization	A05	Mechanical Problem	10/07/2023	10/07/2023
							F12	Serious Injury/ Illness/ Impairment	A05	Mechanical Problem	10/07/2023	10/07/2023
							F1905	Device Revision or Replacement	A05	Mechanical Problem	10/07/2023	10/07/2023
					CT2750	Implantable cardiac electrical stimulators and associated devices	E233001	Chest Pain	A05	Mechanical Problem	10/07/2023	10/07/2023
							F08	Hospitalization or Prolonged Hospitalization	A05	Mechanical Problem	10/07/2023	10/07/2023
							F12	Serious Injury/ Illness/ Impairment	A05	Mechanical Problem	10/07/2023	10/07/2023
							F1905	Device Revision or Replacement	A05	Mechanical Problem	10/07/2023	10/07/2023
					CT2866	Implantable cardiac electrical stimulators	E233001	Chest Pain	A05	Mechanical Problem	10/07/2023	10/07/2023

							F08	Hospitalization or Prolonged Hospitalization	A05	Mechanical Problem	10/07/2023	10/07/2023
							F12	Serious Injury/ Illness/ Impairment	A05	Mechanical Problem	10/07/2023	10/07/2023
							F1905	Device Revision or Replacement	A05	Mechanical Problem	10/07/2023	10/07/2023
					CT2985	Active	E233001	Chest Pain	A05	Mechanical Problem	10/07/2023	10/07/2023
							F08	Hospitalization or Prolonged Hospitalization	A05	Mechanical Problem	10/07/2023	10/07/2023
							F12	Serious Injury/ Illness/ Impairment	A05	Mechanical Problem	10/07/2023	10/07/2023
							F1905	Device Revision or Replacement	A05	Mechanical Problem	10/07/2023	10/07/2023
					CT3001	Implantable metal	E233001	Chest Pain	A05	Mechanical Problem	10/07/2023	10/07/2023
							F08	Hospitalization or Prolonged Hospitalization	A05	Mechanical Problem	10/07/2023	10/07/2023
							F12	Serious Injury/ Illness/ Impairment	A05	Mechanical Problem	10/07/2023	10/07/2023
							F1905	Device Revision or Replacement	A05	Mechanical Problem	10/07/2023	10/07/2023
					CT3002	Implantable synthetic polymer	E233001	Chest Pain	A05	Mechanical Problem	10/07/2023	10/07/2023
							F08	Hospitalization or Prolonged Hospitalization	A05	Mechanical Problem	10/07/2023	10/07/2023
							F12	Serious Injury/ Illness/ Impairment	A05	Mechanical Problem	10/07/2023	10/07/2023
							F1905	Device Revision or Replacement	A05	Mechanical Problem	10/07/2023	10/07/2023
					CT3006	Implantable inorganic material	E233001	Chest Pain	A05	Mechanical Problem	10/07/2023	10/07/2023
							F08	Hospitalization or Prolonged Hospitalization	A05	Mechanical Problem	10/07/2023	10/07/2023
							F12	Serious Injury/ Illness/ Impairment	A05	Mechanical Problem	10/07/2023	10/07/2023
							F1905	Device Revision or Replacement	A05	Mechanical Problem	10/07/2023	10/07/2023
					CT314	Electrically powered	E233001	Chest Pain	A05	Mechanical Problem	10/07/2023	10/07/2023
							F08	Hospitalization or Prolonged Hospitalization	A05	Mechanical Problem	10/07/2023	10/07/2023
							F12	Serious Injury/ Illness/ Impairment	A05	Mechanical Problem	10/07/2023	10/07/2023
							F1905	Device Revision or Replacement	A05	Mechanical Problem	10/07/2023	10/07/2023
					CT321	Long-term surgical invasive	E233001	Chest Pain	A05	Mechanical Problem	10/07/2023	10/07/2023
							F08	Hospitalization or Prolonged Hospitalization	A05	Mechanical Problem	10/07/2023	10/07/2023
							F12	Serious Injury/ Illness/ Impairment	A05	Mechanical Problem	10/07/2023	10/07/2023
							F1905	Device Revision or Replacement	A05	Mechanical Problem	10/07/2023	10/07/2023
					CT449	Generators and associated devices	E233001	Chest Pain	A05	Mechanical Problem	10/07/2023	10/07/2023
							F08	Hospitalization or Prolonged Hospitalization	A05	Mechanical Problem	10/07/2023	10/07/2023
							F12	Serious Injury/ Illness/ Impairment	A05	Mechanical Problem	10/07/2023	10/07/2023
							F1905	Device Revision or Replacement	A05	Mechanical Problem	10/07/2023	10/07/2023
					CT981	Single use	E233001	Chest Pain	A05	Mechanical Problem	10/07/2023	10/07/2023
							F08	Hospitalization or Prolonged Hospitalization	A05	Mechanical Problem	10/07/2023	10/07/2023
							F12	Serious Injury/ Illness/ Impairment	A05	Mechanical Problem	10/07/2023	10/07/2023
							F1905	Device Revision or Replacement	A05	Mechanical Problem	10/07/2023	10/07/2023
					CT983	Surgical invasive	E233001	Chest Pain	A05	Mechanical Problem	10/07/2023	10/07/2023
							F08	Hospitalization or Prolonged Hospitalization	A05	Mechanical Problem	10/07/2023	10/07/2023
							F12	Serious Injury/ Illness/ Impairment	A05	Mechanical Problem	10/07/2023	10/07/2023
							F1905	Device Revision or Replacement	A05	Mechanical Problem	10/07/2023	10/07/2023
13549	2023	All other reportable incidents	47205	Cardiac pulse generator programmer	CT1684	Cardiac pacemakers and associated devices	E2104	Electric Shock	A070909	Oversensing		
									A071302	Inappropriate/Inadequate Shock/Stimulation		
							F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A070909	Oversensing		
									A071302	Inappropriate/Inadequate Shock/Stimulation		
							F23	Unexpected Medical Intervention	A070909	Oversensing		
									A071302	Inappropriate/Inadequate Shock/Stimulation		
13550	2023	All other reportable incidents	47265	Dual-chamber implantable pacemaker, rate-responsive	CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A070102	High Capture Threshold	28/06/2023	28/06/2023
									A0705	Battery Problem	28/06/2023	28/06/2023
									A070504	Premature Discharge of Battery	28/06/2023	28/06/2023
							F08	Hospitalization or Prolonged Hospitalization	A070102	High Capture Threshold	28/06/2023	28/06/2023
									A0705	Battery Problem	28/06/2023	28/06/2023
									A070504	Premature Discharge of Battery	28/06/2023	28/06/2023
							F1905	Device Revision or Replacement	A070102	High Capture Threshold	28/06/2023	28/06/2023
									A0705	Battery Problem	28/06/2023	28/06/2023
									A070504	Premature Discharge of Battery	28/06/2023	28/06/2023
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A070102	High Capture Threshold	28/06/2023	28/06/2023
									A0705	Battery Problem	28/06/2023	28/06/2023
									A070504	Premature Discharge of Battery	28/06/2023	28/06/2023
							F08	Hospitalization or Prolonged Hospitalization	A070102	High Capture Threshold	28/06/2023	28/06/2023
									A0705	Battery Problem	28/06/2023	28/06/2023
									A070504	Premature Discharge of Battery	28/06/2023	28/06/2023
							F1905	Device Revision or Replacement	A070102	High Capture Threshold	28/06/2023	28/06/2023
									A0705	Battery Problem	28/06/2023	28/06/2023
									A070504	Premature Discharge of Battery	28/06/2023	28/06/2023
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A070102	High Capture Threshold	28/06/2023	28/06/2023
									A0705	Battery Problem	28/06/2023	28/06/2023
									A070504	Premature Discharge of Battery	28/06/2023	28/06/2023
							F08	Hospitalization or Prolonged Hospitalization	A070102	High Capture Threshold	28/06/2023	28/06/2023
									A0705	Battery Problem	28/06/2023	28/06/2023
									A070504	Premature Discharge of Battery	28/06/2023	28/06/2023
							F1905	Device Revision or Replacement	A070102	High Capture Threshold	28/06/2023	28/06/2023
									A0705	Battery Problem	28/06/2023	28/06/2023
									A070504	Premature Discharge of Battery	28/06/2023	28/06/2023
13551	2023	All other reportable incidents	35852	Single-chamber implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
13552	2023	All other reportable incidents	47265	Dual-chamber implantable pacemaker, rate-responsive	CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A0908	Incorrect, Inadequate or Imprecise Result or Readings	21/06/2023	21/06/2023
									A160503	Inappropriate or Unexpected Reset	21/06/2023	21/06/2023
							F08	Hospitalization or Prolonged Hospitalization	A0908	Incorrect, Inadequate or Imprecise Result or Readings	21/06/2023	21/06/2023
									A160503	Inappropriate or Unexpected Reset	21/06/2023	21/06/2023
							F1905	Device Revision or Replacement	A0908	Incorrect, Inadequate or Imprecise Result or Readings	21/06/2023	21/06/2023
									A160503	Inappropriate or Unexpected Reset	21/06/2023	21/06/2023
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A0908	Incorrect, Inadequate or Imprecise Result or Readings	21/06/2023	21/06/2023
									A160503	Inappropriate or Unexpected Reset	21/06/2023	21/06/2023

						F08	Hospitalization or Prolonged Hospitalization	A0908	Incorrect, Inadequate or Imprecise Result or Readings	21/06/2023	21/06/2023
								A160503	Inappropriate or Unexpected Reset	21/06/2023	21/06/2023
						F1905	Device Revision or Replacement	A0908	Incorrect, Inadequate or Imprecise Result or Readings	21/06/2023	21/06/2023
								A160503	Inappropriate or Unexpected Reset	21/06/2023	21/06/2023
				CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A0908	Incorrect, Inadequate or Imprecise Result or Readings	21/06/2023	21/06/2023
								A160503	Inappropriate or Unexpected Reset	21/06/2023	21/06/2023
						F08	Hospitalization or Prolonged Hospitalization	A0908	Incorrect, Inadequate or Imprecise Result or Readings	21/06/2023	21/06/2023
								A160503	Inappropriate or Unexpected Reset	21/06/2023	21/06/2023
						F1905	Device Revision or Replacement	A0908	Incorrect, Inadequate or Imprecise Result or Readings	21/06/2023	21/06/2023
								A160503	Inappropriate or Unexpected Reset	21/06/2023	21/06/2023
13553	2023	All other reportable incidents	35852	Single-chamber implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E2104	Electric Shock	A070909	Oversensing		
									A071302	Inappropriate/Inadequate Shock/Stimulation		
									A090801	Signal Artifact/Noise		
						F08	Hospitalization or Prolonged Hospitalization	A070909	Oversensing		
								A071302	Inappropriate/Inadequate Shock/Stimulation		
								A090801	Signal Artifact/Noise		
						F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A070909	Oversensing		
								A071302	Inappropriate/Inadequate Shock/Stimulation		
								A090801	Signal Artifact/Noise		
13554	2023	All other reportable incidents	35223	Endocardial/interventricular septal pacing lead	CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070101	Failure to Capture		
									A070102	High Capture Threshold		
						F26	No Health Consequences or Impact	A070101	Failure to Capture		
								A070102	High Capture Threshold		
13555	2023		47265	Dual-chamber implantable pacemaker, rate-responsive	CT141	Electrodes/Leads and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	21/06/2023	21/06/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	21/06/2023	21/06/2023
					CT145	Cardiovascular devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	21/06/2023	21/06/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	21/06/2023	21/06/2023
					CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	21/06/2023	21/06/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	21/06/2023	21/06/2023
					CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	21/06/2023	21/06/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	21/06/2023	21/06/2023
					CT233	Surgical	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	21/06/2023	21/06/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	21/06/2023	21/06/2023
					CT2406	Implantable	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	21/06/2023	21/06/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	21/06/2023	21/06/2023
					CT2407	Active implantable	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	21/06/2023	21/06/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	21/06/2023	21/06/2023
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	21/06/2023	21/06/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	21/06/2023	21/06/2023
					CT2727	Active implant devices/ancillaries	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	21/06/2023	21/06/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	21/06/2023	21/06/2023
					CT2750	Implantable cardiac electrical stimulators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	21/06/2023	21/06/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	21/06/2023	21/06/2023
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	21/06/2023	21/06/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	21/06/2023	21/06/2023
					CT2985	Active	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	21/06/2023	21/06/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	21/06/2023	21/06/2023
					CT3001	Implantable metal	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	21/06/2023	21/06/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	21/06/2023	21/06/2023
					CT3002	Implantable synthetic polymer	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	21/06/2023	21/06/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	21/06/2023	21/06/2023
					CT3006	Implantable inorganic material	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	21/06/2023	21/06/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	21/06/2023	21/06/2023
					CT314	Electrically powered	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	21/06/2023	21/06/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	21/06/2023	21/06/2023
					CT321	Long-term surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	21/06/2023	21/06/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	21/06/2023	21/06/2023
					CT449	Generators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	21/06/2023	21/06/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	21/06/2023	21/06/2023
					CT981	Single use	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	21/06/2023	21/06/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	21/06/2023	21/06/2023
					CT983	Surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	21/06/2023	21/06/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	21/06/2023	21/06/2023
13556	2023	All other reportable incidents	47263	Cardiac resynchronization therapy implantable pacemaker	CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A1206	Misconnection	07/07/2023	07/07/2023
							F26	No Health Consequences or Impact	A1206	Misconnection	07/07/2023	07/07/2023
					CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A1206	Misconnection	07/07/2023	07/07/2023
							F26	No Health Consequences or Impact	A1206	Misconnection	07/07/2023	07/07/2023
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A1206	Misconnection	07/07/2023	07/07/2023
							F26	No Health Consequences or Impact	A1206	Misconnection	07/07/2023	07/07/2023
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A1206	Misconnection	07/07/2023	07/07/2023
							F26	No Health Consequences or Impact	A1206	Misconnection	07/07/2023	07/07/2023
13557	2023	All other reportable incidents	47265	Dual-chamber implantable pacemaker, rate-responsive	CT141	Electrodes/Leads and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A0908	Incorrect, Inadequate or Imprecise Result or Readings		
							F26	No Health Consequences or Impact	A0908	Incorrect, Inadequate or Imprecise Result or Readings		
					CT145	Cardiovascular devices	E2403	No Clinical Signs, Symptoms or Conditions	A0908	Incorrect, Inadequate or Imprecise Result or Readings		
							F26	No Health Consequences or Impact	A0908	Incorrect, Inadequate or Imprecise Result or Readings		
					CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A0908	Incorrect, Inadequate or Imprecise Result or Readings		
							F26	No Health Consequences or Impact	A0908	Incorrect, Inadequate or Imprecise Result or Readings		
					CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A0908	Incorrect, Inadequate or Imprecise Result or Readings		

							F26	No Health Consequences or Impact	A0908	Incorrect, Inadequate or Imprecise Result or Readings		
					CT233	Surgical	E2403	No Clinical Signs, Symptoms or Conditions	A0908	Incorrect, Inadequate or Imprecise Result or Readings		
							F26	No Health Consequences or Impact	A0908	Incorrect, Inadequate or Imprecise Result or Readings		
					CT2406	Implantable	E2403	No Clinical Signs, Symptoms or Conditions	A0908	Incorrect, Inadequate or Imprecise Result or Readings		
							F26	No Health Consequences or Impact	A0908	Incorrect, Inadequate or Imprecise Result or Readings		
					CT2407	Active implantable	E2403	No Clinical Signs, Symptoms or Conditions	A0908	Incorrect, Inadequate or Imprecise Result or Readings		
							F26	No Health Consequences or Impact	A0908	Incorrect, Inadequate or Imprecise Result or Readings		
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A0908	Incorrect, Inadequate or Imprecise Result or Readings		
							F26	No Health Consequences or Impact	A0908	Incorrect, Inadequate or Imprecise Result or Readings		
					CT2727	Active implant devices/ancillaries	E2403	No Clinical Signs, Symptoms or Conditions	A0908	Incorrect, Inadequate or Imprecise Result or Readings		
							F26	No Health Consequences or Impact	A0908	Incorrect, Inadequate or Imprecise Result or Readings		
					CT2750	Implantable cardiac electrical stimulators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A0908	Incorrect, Inadequate or Imprecise Result or Readings		
							F26	No Health Consequences or Impact	A0908	Incorrect, Inadequate or Imprecise Result or Readings		
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A0908	Incorrect, Inadequate or Imprecise Result or Readings		
							F26	No Health Consequences or Impact	A0908	Incorrect, Inadequate or Imprecise Result or Readings		
					CT2985	Active	E2403	No Clinical Signs, Symptoms or Conditions	A0908	Incorrect, Inadequate or Imprecise Result or Readings		
							F26	No Health Consequences or Impact	A0908	Incorrect, Inadequate or Imprecise Result or Readings		
					CT3001	Implantable metal	E2403	No Clinical Signs, Symptoms or Conditions	A0908	Incorrect, Inadequate or Imprecise Result or Readings		
							F26	No Health Consequences or Impact	A0908	Incorrect, Inadequate or Imprecise Result or Readings		
					CT3002	Implantable synthetic polymer	E2403	No Clinical Signs, Symptoms or Conditions	A0908	Incorrect, Inadequate or Imprecise Result or Readings		
							F26	No Health Consequences or Impact	A0908	Incorrect, Inadequate or Imprecise Result or Readings		
					CT3006	Implantable inorganic material	E2403	No Clinical Signs, Symptoms or Conditions	A0908	Incorrect, Inadequate or Imprecise Result or Readings		
							F26	No Health Consequences or Impact	A0908	Incorrect, Inadequate or Imprecise Result or Readings		
					CT314	Electrically powered	E2403	No Clinical Signs, Symptoms or Conditions	A0908	Incorrect, Inadequate or Imprecise Result or Readings		
							F26	No Health Consequences or Impact	A0908	Incorrect, Inadequate or Imprecise Result or Readings		
					CT321	Long-term surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A0908	Incorrect, Inadequate or Imprecise Result or Readings		
							F26	No Health Consequences or Impact	A0908	Incorrect, Inadequate or Imprecise Result or Readings		
					CT449	Generators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A0908	Incorrect, Inadequate or Imprecise Result or Readings		
							F26	No Health Consequences or Impact	A0908	Incorrect, Inadequate or Imprecise Result or Readings		
					CT981	Single use	E2403	No Clinical Signs, Symptoms or Conditions	A0908	Incorrect, Inadequate or Imprecise Result or Readings		
							F26	No Health Consequences or Impact	A0908	Incorrect, Inadequate or Imprecise Result or Readings		
					CT983	Surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A0908	Incorrect, Inadequate or Imprecise Result or Readings		
							F26	No Health Consequences or Impact	A0908	Incorrect, Inadequate or Imprecise Result or Readings		
13558	2023	All other reportable incidents	36241	Pacemaker/defibrillator lead	CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A072201	High Impedance		
							F2203	Imaging Required	A072201	High Impedance		
							F26	No Health Consequences or Impact	A072201	High Impedance		
13559	2023	All other reportable incidents	17882	Physiologic-monitoring defibrillation system	CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070803	Failure to Power Up		
							F27	Problem Identified During Non-Clinical Procedure	A070803	Failure to Power Up		
13560	2023	All other reportable incidents	47265	Dual-chamber implantable pacemaker, rate-responsive	CT141	Electrodes/Leads and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A1605	Reset Problem		
							F26	No Health Consequences or Impact	A1605	Reset Problem		
					CT145	Cardiovascular devices	E2403	No Clinical Signs, Symptoms or Conditions	A1605	Reset Problem		
							F26	No Health Consequences or Impact	A1605	Reset Problem		
					CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A1605	Reset Problem		
							F26	No Health Consequences or Impact	A1605	Reset Problem		
					CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A1605	Reset Problem		
							F26	No Health Consequences or Impact	A1605	Reset Problem		
					CT233	Surgical	E2403	No Clinical Signs, Symptoms or Conditions	A1605	Reset Problem		
							F26	No Health Consequences or Impact	A1605	Reset Problem		
					CT2406	Implantable	E2403	No Clinical Signs, Symptoms or Conditions	A1605	Reset Problem		
							F26	No Health Consequences or Impact	A1605	Reset Problem		
					CT2407	Active implantable	E2403	No Clinical Signs, Symptoms or Conditions	A1605	Reset Problem		
							F26	No Health Consequences or Impact	A1605	Reset Problem		
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A1605	Reset Problem		
							F26	No Health Consequences or Impact	A1605	Reset Problem		
					CT2727	Active implant devices/ancillaries	E2403	No Clinical Signs, Symptoms or Conditions	A1605	Reset Problem		
							F26	No Health Consequences or Impact	A1605	Reset Problem		
					CT2750	Implantable cardiac electrical stimulators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A1605	Reset Problem		
							F26	No Health Consequences or Impact	A1605	Reset Problem		
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A1605	Reset Problem		
							F26	No Health Consequences or Impact	A1605	Reset Problem		
					CT2985	Active	E2403	No Clinical Signs, Symptoms or Conditions	A1605	Reset Problem		
							F26	No Health Consequences or Impact	A1605	Reset Problem		
					CT3001	Implantable metal	E2403	No Clinical Signs, Symptoms or Conditions	A1605	Reset Problem		

							F26	No Health Consequences or Impact	A1605	Reset Problem			
					CT3002	Implantable synthetic polymer	E2403	No Clinical Signs, Symptoms or Conditions	A1605	Reset Problem			
							F26	No Health Consequences or Impact	A1605	Reset Problem			
					CT3006	Implantable inorganic material	E2403	No Clinical Signs, Symptoms or Conditions	A1605	Reset Problem			
							F26	No Health Consequences or Impact	A1605	Reset Problem			
					CT314	Electrically powered	E2403	No Clinical Signs, Symptoms or Conditions	A1605	Reset Problem			
							F26	No Health Consequences or Impact	A1605	Reset Problem			
					CT321	Long-term surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A1605	Reset Problem			
							F26	No Health Consequences or Impact	A1605	Reset Problem			
					CT449	Generators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A1605	Reset Problem			
							F26	No Health Consequences or Impact	A1605	Reset Problem			
					CT981	Single use	E2403	No Clinical Signs, Symptoms or Conditions	A1605	Reset Problem			
							F26	No Health Consequences or Impact	A1605	Reset Problem			
					CT983	Surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A1605	Reset Problem			
							F26	No Health Consequences or Impact	A1605	Reset Problem			
13561	2023	All other reportable incidents	47270	Cardiac resynchronization therapy implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A070909	Oversensing			
									A072001	Electromagnetic Interference			
									A090801	Signal Artifact/Noise			
							F26	No Health Consequences or Impact	A070909	Oversensing			
									A072001	Electromagnetic Interference			
									A090801	Signal Artifact/Noise			
13562	2023	All other reportable incidents	35223	Endocardial/interventricular septal pacing lead	CT1684	Cardiac pacemakers and associated devices	E011902	Presyncope	A070910	Undersensing			
							E2403	No Clinical Signs, Symptoms or Conditions	A070910	Undersensing			
							F12	Serious Injury/ Illness/ Impairment	A070910	Undersensing			
							F23	Unexpected Medical Intervention	A070910	Undersensing			
13563	2023	All other reportable incidents	37265	Dual-chamber implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E060109	Tachycardia	A0712	Pacing Problem			
							F12	Serious Injury/ Illness/ Impairment	A0712	Pacing Problem			
							F23	Unexpected Medical Intervention	A0712	Pacing Problem			
13564	2023	All other reportable incidents	35223	Endocardial/interventricular septal pacing lead	CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A0709	Device Sensing Problem			
									A090801	Signal Artifact/Noise			
							F12	Serious Injury/ Illness/ Impairment	A0709	Device Sensing Problem			
									A090801	Signal Artifact/Noise			
							F23	Unexpected Medical Intervention	A0709	Device Sensing Problem			
									A090801	Signal Artifact/Noise			
13565	2023	All other reportable incidents	36241	Pacemaker/defibrillator lead	CT1684	Cardiac pacemakers and associated devices	E1906	Unspecified Infection	A24	No Apparent Device Problem	12/04/2023	12/04/2023	
							F08	Hospitalization or Prolonged Hospitalization	A24	No Apparent Device Problem	12/04/2023	12/04/2023	
							F1903	Device Explantation	A24	No Apparent Device Problem	12/04/2023	12/04/2023	
13566	2023	All other reportable incidents	47265	Dual-chamber implantable pacemaker, rate-responsive	CT141	Electrodes/Leads and associated devices	E0601	Arrhythmia	A070909	Oversensing	05/07/2023	05/07/2023	
									A071102	Failure to Interrogate	05/07/2023	05/07/2023	
									A071204	No Pacing	05/07/2023	05/07/2023	
									A0723	Unexpected Mode Switch	05/07/2023	05/07/2023	
							F1901	Additional Surgery	A070909	Oversensing	05/07/2023	05/07/2023	
									A071102	Failure to Interrogate	05/07/2023	05/07/2023	
									A071204	No Pacing	05/07/2023	05/07/2023	
									A0723	Unexpected Mode Switch	05/07/2023	05/07/2023	
					CT145	Cardiovascular devices	E0601	Arrhythmia	A070909	Oversensing	05/07/2023	05/07/2023	
									A071102	Failure to Interrogate	05/07/2023	05/07/2023	
									A071204	No Pacing	05/07/2023	05/07/2023	
									A0723	Unexpected Mode Switch	05/07/2023	05/07/2023	
							F1901	Additional Surgery	A070909	Oversensing	05/07/2023	05/07/2023	
									A071102	Failure to Interrogate	05/07/2023	05/07/2023	
									A071204	No Pacing	05/07/2023	05/07/2023	
									A0723	Unexpected Mode Switch	05/07/2023	05/07/2023	
					CT1684	Cardiac pacemakers and associated devices	E0601	Arrhythmia	A070909	Oversensing	05/07/2023	05/07/2023	
									A071102	Failure to Interrogate	05/07/2023	05/07/2023	
									A071204	No Pacing	05/07/2023	05/07/2023	
									A0723	Unexpected Mode Switch	05/07/2023	05/07/2023	
							F1901	Additional Surgery	A070909	Oversensing	05/07/2023	05/07/2023	
									A071102	Failure to Interrogate	05/07/2023	05/07/2023	
									A071204	No Pacing	05/07/2023	05/07/2023	
									A0723	Unexpected Mode Switch	05/07/2023	05/07/2023	
					CT1685	Cardiac pacemakers	E0601	Arrhythmia	A070909	Oversensing	05/07/2023	05/07/2023	
									A071102	Failure to Interrogate	05/07/2023	05/07/2023	
									A071204	No Pacing	05/07/2023	05/07/2023	
									A0723	Unexpected Mode Switch	05/07/2023	05/07/2023	
							F1901	Additional Surgery	A070909	Oversensing	05/07/2023	05/07/2023	
									A071102	Failure to Interrogate	05/07/2023	05/07/2023	
									A071204	No Pacing	05/07/2023	05/07/2023	
									A0723	Unexpected Mode Switch	05/07/2023	05/07/2023	
					CT233	Surgical	E0601	Arrhythmia	A070909	Oversensing	05/07/2023	05/07/2023	
									A071102	Failure to Interrogate	05/07/2023	05/07/2023	
									A071204	No Pacing	05/07/2023	05/07/2023	
									A0723	Unexpected Mode Switch	05/07/2023	05/07/2023	
							F1901	Additional Surgery	A070909	Oversensing	05/07/2023	05/07/2023	
									A071102	Failure to Interrogate	05/07/2023	05/07/2023	
									A071204	No Pacing	05/07/2023	05/07/2023	
									A0723	Unexpected Mode Switch	05/07/2023	05/07/2023	
					CT2406	Implantable	E0601	Arrhythmia	A070909	Oversensing	05/07/2023	05/07/2023	
									A071102	Failure to Interrogate	05/07/2023	05/07/2023	
									A071204	No Pacing	05/07/2023	05/07/2023	
									A0723	Unexpected Mode Switch	05/07/2023	05/07/2023	
							F1901	Additional Surgery	A070909	Oversensing	05/07/2023	05/07/2023	
									A071102	Failure to Interrogate	05/07/2023	05/07/2023	
									A071204	No Pacing	05/07/2023	05/07/2023	
									A0723	Unexpected Mode Switch	05/07/2023	05/07/2023	
					CT2407	Active implantable	E0601	Arrhythmia	A070909	Oversensing	05/07/2023	05/07/2023	
									A071102	Failure to Interrogate	05/07/2023	05/07/2023	
									A071204	No Pacing	05/07/2023	05/07/2023	

								A0723	Unexpected Mode Switch	05/07/2023	05/07/2023
							F1901	Additional Surgery	A070909	Oversensing	05/07/2023	05/07/2023
									A071102	Failure to Interrogate	05/07/2023	05/07/2023
									A071204	No Pacing	05/07/2023	05/07/2023
									A0723	Unexpected Mode Switch	05/07/2023	05/07/2023
					CT257	Implantable cardiac pacemakers	E0601	Arrhythmia	A070909	Oversensing	05/07/2023	05/07/2023
									A071102	Failure to Interrogate	05/07/2023	05/07/2023
									A071204	No Pacing	05/07/2023	05/07/2023
									A0723	Unexpected Mode Switch	05/07/2023	05/07/2023
							F1901	Additional Surgery	A070909	Oversensing	05/07/2023	05/07/2023
									A071102	Failure to Interrogate	05/07/2023	05/07/2023
									A071204	No Pacing	05/07/2023	05/07/2023
									A0723	Unexpected Mode Switch	05/07/2023	05/07/2023
					CT2727	Active implant devices/ancillaries	E0601	Arrhythmia	A070909	Oversensing	05/07/2023	05/07/2023
									A071102	Failure to Interrogate	05/07/2023	05/07/2023
									A071204	No Pacing	05/07/2023	05/07/2023
									A0723	Unexpected Mode Switch	05/07/2023	05/07/2023
							F1901	Additional Surgery	A070909	Oversensing	05/07/2023	05/07/2023
									A071102	Failure to Interrogate	05/07/2023	05/07/2023
									A071204	No Pacing	05/07/2023	05/07/2023
									A0723	Unexpected Mode Switch	05/07/2023	05/07/2023
					CT2750	Implantable cardiac electrical stimulators and associated devices	E0601	Arrhythmia	A070909	Oversensing	05/07/2023	05/07/2023
									A071102	Failure to Interrogate	05/07/2023	05/07/2023
									A071204	No Pacing	05/07/2023	05/07/2023
									A0723	Unexpected Mode Switch	05/07/2023	05/07/2023
							F1901	Additional Surgery	A070909	Oversensing	05/07/2023	05/07/2023
									A071102	Failure to Interrogate	05/07/2023	05/07/2023
									A071204	No Pacing	05/07/2023	05/07/2023
									A0723	Unexpected Mode Switch	05/07/2023	05/07/2023
					CT2866	Implantable cardiac electrical stimulators	E0601	Arrhythmia	A070909	Oversensing	05/07/2023	05/07/2023
									A071102	Failure to Interrogate	05/07/2023	05/07/2023
									A071204	No Pacing	05/07/2023	05/07/2023
									A0723	Unexpected Mode Switch	05/07/2023	05/07/2023
							F1901	Additional Surgery	A070909	Oversensing	05/07/2023	05/07/2023
									A071102	Failure to Interrogate	05/07/2023	05/07/2023
									A071204	No Pacing	05/07/2023	05/07/2023
									A0723	Unexpected Mode Switch	05/07/2023	05/07/2023
					CT2985	Active	E0601	Arrhythmia	A070909	Oversensing	05/07/2023	05/07/2023
									A071102	Failure to Interrogate	05/07/2023	05/07/2023
									A071204	No Pacing	05/07/2023	05/07/2023
									A0723	Unexpected Mode Switch	05/07/2023	05/07/2023
							F1901	Additional Surgery	A070909	Oversensing	05/07/2023	05/07/2023
									A071102	Failure to Interrogate	05/07/2023	05/07/2023
									A071204	No Pacing	05/07/2023	05/07/2023
									A0723	Unexpected Mode Switch	05/07/2023	05/07/2023
					CT3001	Implantable metal	E0601	Arrhythmia	A070909	Oversensing	05/07/2023	05/07/2023
									A071102	Failure to Interrogate	05/07/2023	05/07/2023
									A071204	No Pacing	05/07/2023	05/07/2023
									A0723	Unexpected Mode Switch	05/07/2023	05/07/2023
							F1901	Additional Surgery	A070909	Oversensing	05/07/2023	05/07/2023
									A071102	Failure to Interrogate	05/07/2023	05/07/2023
									A071204	No Pacing	05/07/2023	05/07/2023
									A0723	Unexpected Mode Switch	05/07/2023	05/07/2023
					CT3002	Implantable synthetic polymer	E0601	Arrhythmia	A070909	Oversensing	05/07/2023	05/07/2023
									A071102	Failure to Interrogate	05/07/2023	05/07/2023
									A071204	No Pacing	05/07/2023	05/07/2023
									A0723	Unexpected Mode Switch	05/07/2023	05/07/2023
							F1901	Additional Surgery	A070909	Oversensing	05/07/2023	05/07/2023
									A071102	Failure to Interrogate	05/07/2023	05/07/2023
									A071204	No Pacing	05/07/2023	05/07/2023
									A0723	Unexpected Mode Switch	05/07/2023	05/07/2023
					CT3006	Implantable inorganic material	E0601	Arrhythmia	A070909	Oversensing	05/07/2023	05/07/2023
									A071102	Failure to Interrogate	05/07/2023	05/07/2023
									A071204	No Pacing	05/07/2023	05/07/2023
									A0723	Unexpected Mode Switch	05/07/2023	05/07/2023
							F1901	Additional Surgery	A070909	Oversensing	05/07/2023	05/07/2023
									A071102	Failure to Interrogate	05/07/2023	05/07/2023
									A071204	No Pacing	05/07/2023	05/07/2023
									A0723	Unexpected Mode Switch	05/07/2023	05/07/2023
					CT314	Electrically powered	E0601	Arrhythmia	A070909	Oversensing	05/07/2023	05/07/2023
									A071102	Failure to Interrogate	05/07/2023	05/07/2023
									A071204	No Pacing	05/07/2023	05/07/2023
									A0723	Unexpected Mode Switch	05/07/2023	05/07/2023
							F1901	Additional Surgery	A070909	Oversensing	05/07/2023	05/07/2023
									A071102	Failure to Interrogate	05/07/2023	05/07/2023
									A071204	No Pacing	05/07/2023	05/07/2023
									A0723	Unexpected Mode Switch	05/07/2023	05/07/2023
					CT321	Long-term surgical invasive	E0601	Arrhythmia	A070909	Oversensing	05/07/2023	05/07/2023
									A071102	Failure to Interrogate	05/07/2023	05/07/2023
									A071204	No Pacing	05/07/2023	05/07/2023
									A0723	Unexpected Mode Switch	05/07/2023	05/07/2023
							F1901	Additional Surgery	A070909	Oversensing	05/07/2023	05/07/2023
									A071102	Failure to Interrogate	05/07/2023	05/07/2023
									A071204	No Pacing	05/07/2023	05/07/2023
									A0723	Unexpected Mode Switch	05/07/2023	05/07/2023
					CT449	Generators and associated devices	E0601	Arrhythmia	A070909	Oversensing	05/07/2023	05/07/2023
									A071102	Failure to Interrogate	05/07/2023	05/07/2023
									A071204	No Pacing	05/07/2023	05/07/2023

									A0723	Unexpected Mode Switch	05/07/2023	05/07/2023
							F1901	Additional Surgery	A070909	Oversensing	05/07/2023	05/07/2023
									A071102	Failure to Interrogate	05/07/2023	05/07/2023
									A071204	No Pacing	05/07/2023	05/07/2023
									A0723	Unexpected Mode Switch	05/07/2023	05/07/2023
					CT981	Single use	E0601	Arrhythmia	A070909	Oversensing	05/07/2023	05/07/2023
									A071102	Failure to Interrogate	05/07/2023	05/07/2023
									A071204	No Pacing	05/07/2023	05/07/2023
									A0723	Unexpected Mode Switch	05/07/2023	05/07/2023
							F1901	Additional Surgery	A070909	Oversensing	05/07/2023	05/07/2023
									A071102	Failure to Interrogate	05/07/2023	05/07/2023
									A071204	No Pacing	05/07/2023	05/07/2023
									A0723	Unexpected Mode Switch	05/07/2023	05/07/2023
					CT983	Surgical invasive	E0601	Arrhythmia	A070909	Oversensing	05/07/2023	05/07/2023
									A071102	Failure to Interrogate	05/07/2023	05/07/2023
									A071204	No Pacing	05/07/2023	05/07/2023
									A0723	Unexpected Mode Switch	05/07/2023	05/07/2023
							F1901	Additional Surgery	A070909	Oversensing	05/07/2023	05/07/2023
									A071102	Failure to Interrogate	05/07/2023	05/07/2023
									A071204	No Pacing	05/07/2023	05/07/2023
									A0723	Unexpected Mode Switch	05/07/2023	05/07/2023
13567	2023	All other reportable incidents	17882	Physiologic-monitoring defibrillation system	CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A090204	Image Display Error/Artifact		
							F27	Problem Identified During Non-Clinical Procedure	A090204	Image Display Error/Artifact		
13568	2023	All other reportable incidents	47265	Dual-chamber implantable pacemaker, rate-responsive	CT141	Electrodes/Leads and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	04/04/2023	04/04/2023
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	04/04/2023	04/04/2023
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	04/04/2023	04/04/2023
					CT145	Cardiovascular devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	04/04/2023	04/04/2023
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	04/04/2023	04/04/2023
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	04/04/2023	04/04/2023
					CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	04/04/2023	04/04/2023
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	04/04/2023	04/04/2023
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	04/04/2023	04/04/2023
					CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	04/04/2023	04/04/2023
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	04/04/2023	04/04/2023
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	04/04/2023	04/04/2023
					CT233	Surgical	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	04/04/2023	04/04/2023
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	04/04/2023	04/04/2023
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	04/04/2023	04/04/2023
					CT2406	Implantable	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	04/04/2023	04/04/2023
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	04/04/2023	04/04/2023
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	04/04/2023	04/04/2023
					CT2407	Active implantable	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	04/04/2023	04/04/2023
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	04/04/2023	04/04/2023
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	04/04/2023	04/04/2023
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	04/04/2023	04/04/2023
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	04/04/2023	04/04/2023
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	04/04/2023	04/04/2023
					CT2727	Active implant devices/ancillaries	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	04/04/2023	04/04/2023
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	04/04/2023	04/04/2023
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	04/04/2023	04/04/2023
					CT2750	Implantable cardiac electrical stimulators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	04/04/2023	04/04/2023
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	04/04/2023	04/04/2023
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	04/04/2023	04/04/2023
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	04/04/2023	04/04/2023
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	04/04/2023	04/04/2023
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	04/04/2023	04/04/2023
					CT2985	Active	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	04/04/2023	04/04/2023
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	04/04/2023	04/04/2023
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	04/04/2023	04/04/2023
					CT3001	Implantable metal	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	04/04/2023	04/04/2023
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	04/04/2023	04/04/2023
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	04/04/2023	04/04/2023
					CT3002	Implantable synthetic polymer	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	04/04/2023	04/04/2023
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	04/04/2023	04/04/2023
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	04/04/2023	04/04/2023
					CT3006	Implantable inorganic material	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	04/04/2023	04/04/2023
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	04/04/2023	04/04/2023
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	04/04/2023	04/04/2023
					CT314	Electrically powered	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	04/04/2023	04/04/2023
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	04/04/2023	04/04/2023
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	04/04/2023	04/04/2023
					CT321	Long-term surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	04/04/2023	04/04/2023
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	04/04/2023	04/04/2023
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	04/04/2023	04/04/2023
					CT449	Generators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	04/04/2023	04/04/2023
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	04/04/2023	04/04/2023
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	04/04/2023	04/04/2023
					CT981	Single use	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	04/04/2023	04/04/2023
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	04/04/2023	04/04/2023
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	04/04/2023	04/04/2023
					CT983	Surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	04/04/2023	04/04/2023
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	04/04/2023	04/04/2023
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	04/04/2023	04/04/2023
13569	2023	All other reportable incidents	47265	Dual-chamber implantable pacemaker, rate-responsive	CT141	Electrodes/Leads and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A0908	Incorrect, Inadequate or Imprecise Result or Readings	30/06/2023	30/06/2023
							F08	Hospitalization or Prolonged Hospitalization	A0908	Incorrect, Inadequate or Imprecise Result or Readings	30/06/2023	30/06/2023

						F1905	Device Revision or Replacement	A0908	Incorrect, Inadequate or Imprecise Result or Readings	30/06/2023	30/06/2023
					CT145	E2403	No Clinical Signs, Symptoms or Conditions	A0908	Incorrect, Inadequate or Imprecise Result or Readings	30/06/2023	30/06/2023
						F08	Hospitalization or Prolonged Hospitalization	A0908	Incorrect, Inadequate or Imprecise Result or Readings	30/06/2023	30/06/2023
						F1905	Device Revision or Replacement	A0908	Incorrect, Inadequate or Imprecise Result or Readings	30/06/2023	30/06/2023
					CT1684	E2403	No Clinical Signs, Symptoms or Conditions	A0908	Incorrect, Inadequate or Imprecise Result or Readings	30/06/2023	30/06/2023
						F08	Hospitalization or Prolonged Hospitalization	A0908	Incorrect, Inadequate or Imprecise Result or Readings	30/06/2023	30/06/2023
						F1905	Device Revision or Replacement	A0908	Incorrect, Inadequate or Imprecise Result or Readings	30/06/2023	30/06/2023
					CT1685	E2403	No Clinical Signs, Symptoms or Conditions	A0908	Incorrect, Inadequate or Imprecise Result or Readings	30/06/2023	30/06/2023
						F08	Hospitalization or Prolonged Hospitalization	A0908	Incorrect, Inadequate or Imprecise Result or Readings	30/06/2023	30/06/2023
						F1905	Device Revision or Replacement	A0908	Incorrect, Inadequate or Imprecise Result or Readings	30/06/2023	30/06/2023
					CT233	E2403	No Clinical Signs, Symptoms or Conditions	A0908	Incorrect, Inadequate or Imprecise Result or Readings	30/06/2023	30/06/2023
						F08	Hospitalization or Prolonged Hospitalization	A0908	Incorrect, Inadequate or Imprecise Result or Readings	30/06/2023	30/06/2023
						F1905	Device Revision or Replacement	A0908	Incorrect, Inadequate or Imprecise Result or Readings	30/06/2023	30/06/2023
					CT2406	E2403	No Clinical Signs, Symptoms or Conditions	A0908	Incorrect, Inadequate or Imprecise Result or Readings	30/06/2023	30/06/2023
						F08	Hospitalization or Prolonged Hospitalization	A0908	Incorrect, Inadequate or Imprecise Result or Readings	30/06/2023	30/06/2023
						F1905	Device Revision or Replacement	A0908	Incorrect, Inadequate or Imprecise Result or Readings	30/06/2023	30/06/2023
					CT2407	E2403	No Clinical Signs, Symptoms or Conditions	A0908	Incorrect, Inadequate or Imprecise Result or Readings	30/06/2023	30/06/2023
						F08	Hospitalization or Prolonged Hospitalization	A0908	Incorrect, Inadequate or Imprecise Result or Readings	30/06/2023	30/06/2023
						F1905	Device Revision or Replacement	A0908	Incorrect, Inadequate or Imprecise Result or Readings	30/06/2023	30/06/2023
					CT257	E2403	No Clinical Signs, Symptoms or Conditions	A0908	Incorrect, Inadequate or Imprecise Result or Readings	30/06/2023	30/06/2023
						F08	Hospitalization or Prolonged Hospitalization	A0908	Incorrect, Inadequate or Imprecise Result or Readings	30/06/2023	30/06/2023
						F1905	Device Revision or Replacement	A0908	Incorrect, Inadequate or Imprecise Result or Readings	30/06/2023	30/06/2023
					CT2727	E2403	No Clinical Signs, Symptoms or Conditions	A0908	Incorrect, Inadequate or Imprecise Result or Readings	30/06/2023	30/06/2023
						F08	Hospitalization or Prolonged Hospitalization	A0908	Incorrect, Inadequate or Imprecise Result or Readings	30/06/2023	30/06/2023
						F1905	Device Revision or Replacement	A0908	Incorrect, Inadequate or Imprecise Result or Readings	30/06/2023	30/06/2023
					CT2750	E2403	No Clinical Signs, Symptoms or Conditions	A0908	Incorrect, Inadequate or Imprecise Result or Readings	30/06/2023	30/06/2023
						F08	Hospitalization or Prolonged Hospitalization	A0908	Incorrect, Inadequate or Imprecise Result or Readings	30/06/2023	30/06/2023
						F1905	Device Revision or Replacement	A0908	Incorrect, Inadequate or Imprecise Result or Readings	30/06/2023	30/06/2023
					CT2866	E2403	No Clinical Signs, Symptoms or Conditions	A0908	Incorrect, Inadequate or Imprecise Result or Readings	30/06/2023	30/06/2023
						F08	Hospitalization or Prolonged Hospitalization	A0908	Incorrect, Inadequate or Imprecise Result or Readings	30/06/2023	30/06/2023
						F1905	Device Revision or Replacement	A0908	Incorrect, Inadequate or Imprecise Result or Readings	30/06/2023	30/06/2023
					CT2985	E2403	No Clinical Signs, Symptoms or Conditions	A0908	Incorrect, Inadequate or Imprecise Result or Readings	30/06/2023	30/06/2023
						F08	Hospitalization or Prolonged Hospitalization	A0908	Incorrect, Inadequate or Imprecise Result or Readings	30/06/2023	30/06/2023
						F1905	Device Revision or Replacement	A0908	Incorrect, Inadequate or Imprecise Result or Readings	30/06/2023	30/06/2023
					CT3001	E2403	No Clinical Signs, Symptoms or Conditions	A0908	Incorrect, Inadequate or Imprecise Result or Readings	30/06/2023	30/06/2023
						F08	Hospitalization or Prolonged Hospitalization	A0908	Incorrect, Inadequate or Imprecise Result or Readings	30/06/2023	30/06/2023
						F1905	Device Revision or Replacement	A0908	Incorrect, Inadequate or Imprecise Result or Readings	30/06/2023	30/06/2023
					CT3002	E2403	No Clinical Signs, Symptoms or Conditions	A0908	Incorrect, Inadequate or Imprecise Result or Readings	30/06/2023	30/06/2023
						F08	Hospitalization or Prolonged Hospitalization	A0908	Incorrect, Inadequate or Imprecise Result or Readings	30/06/2023	30/06/2023
						F1905	Device Revision or Replacement	A0908	Incorrect, Inadequate or Imprecise Result or Readings	30/06/2023	30/06/2023
					CT3006	E2403	No Clinical Signs, Symptoms or Conditions	A0908	Incorrect, Inadequate or Imprecise Result or Readings	30/06/2023	30/06/2023
						F08	Hospitalization or Prolonged Hospitalization	A0908	Incorrect, Inadequate or Imprecise Result or Readings	30/06/2023	30/06/2023
						F1905	Device Revision or Replacement	A0908	Incorrect, Inadequate or Imprecise Result or Readings	30/06/2023	30/06/2023
					CT314	E2403	No Clinical Signs, Symptoms or Conditions	A0908	Incorrect, Inadequate or Imprecise Result or Readings	30/06/2023	30/06/2023
						F08	Hospitalization or Prolonged Hospitalization	A0908	Incorrect, Inadequate or Imprecise Result or Readings	30/06/2023	30/06/2023
						F1905	Device Revision or Replacement	A0908	Incorrect, Inadequate or Imprecise Result or Readings	30/06/2023	30/06/2023
					CT321	E2403	No Clinical Signs, Symptoms or Conditions	A0908	Incorrect, Inadequate or Imprecise Result or Readings	30/06/2023	30/06/2023
						F08	Hospitalization or Prolonged Hospitalization	A0908	Incorrect, Inadequate or Imprecise Result or Readings	30/06/2023	30/06/2023
						F1905	Device Revision or Replacement	A0908	Incorrect, Inadequate or Imprecise Result or Readings	30/06/2023	30/06/2023
					CT449	E2403	No Clinical Signs, Symptoms or Conditions	A0908	Incorrect, Inadequate or Imprecise Result or Readings	30/06/2023	30/06/2023

							F08	Hospitalization or Prolonged Hospitalization	A0908	Incorrect, Inadequate or Imprecise Result or Readings	30/06/2023	30/06/2023
							F1905	Device Revision or Replacement	A0908	Incorrect, Inadequate or Imprecise Result or Readings	30/06/2023	30/06/2023
					CT981	Single use	E2403	No Clinical Signs, Symptoms or Conditions	A0908	Incorrect, Inadequate or Imprecise Result or Readings	30/06/2023	30/06/2023
							F08	Hospitalization or Prolonged Hospitalization	A0908	Incorrect, Inadequate or Imprecise Result or Readings	30/06/2023	30/06/2023
							F1905	Device Revision or Replacement	A0908	Incorrect, Inadequate or Imprecise Result or Readings	30/06/2023	30/06/2023
					CT983	Surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A0908	Incorrect, Inadequate or Imprecise Result or Readings	30/06/2023	30/06/2023
							F08	Hospitalization or Prolonged Hospitalization	A0908	Incorrect, Inadequate or Imprecise Result or Readings	30/06/2023	30/06/2023
							F1905	Device Revision or Replacement	A0908	Incorrect, Inadequate or Imprecise Result or Readings	30/06/2023	30/06/2023
13570	2023	All other reportable incidents	47265	Dual-chamber implantable pacemaker, rate-responsive	CT141	Electrodes/Leads and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT145	Cardiovascular devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT233	Surgical	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT2406	Implantable	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT2407	Active implantable	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT2727	Active implant devices/ancillaries	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT2750	Implantable cardiac electrical stimulators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT2985	Active	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT3001	Implantable metal	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT3002	Implantable synthetic polymer	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT3006	Implantable inorganic material	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT314	Electrically powered	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT321	Long-term surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT449	Generators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT981	Single use	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT983	Surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
13571	2023	All other reportable incidents	36241	Pacemaker/defibrillator lead	CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070102	High Capture Threshold		
									A070909	Oversensing		
									A090801	Signal Artifact/Noise		
									A1301	Failure to Read Input Signal		
							F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A070102	High Capture Threshold		
									A070909	Oversensing		
									A090801	Signal Artifact/Noise		
									A1301	Failure to Read Input Signal		
13572	2023	All other reportable incidents	47267	Single-chamber implantable pacemaker, rate-responsive	CT141	Electrodes/Leads and associated devices	E2103	Device Overstimulation of Tissue	A070101	Failure to Capture		
									A070102	High Capture Threshold		
									A071209	Extracardiac Stimulation		
									A1301	Failure to Read Input Signal		
							F26	No Health Consequences or Impact	A070101	Failure to Capture		
									A070102	High Capture Threshold		
									A071209	Extracardiac Stimulation		
									A1301	Failure to Read Input Signal		
					CT145	Cardiovascular devices	E2103	Device Overstimulation of Tissue	A070101	Failure to Capture		
									A070102	High Capture Threshold		
									A071209	Extracardiac Stimulation		
									A1301	Failure to Read Input Signal		
							F26	No Health Consequences or Impact	A070101	Failure to Capture		
									A070102	High Capture Threshold		
									A071209	Extracardiac Stimulation		
									A1301	Failure to Read Input Signal		
					CT1684	Cardiac pacemakers and associated devices	E2103	Device Overstimulation of Tissue	A070101	Failure to Capture		
									A070102	High Capture Threshold		
									A071209	Extracardiac Stimulation		
									A1301	Failure to Read Input Signal		
							F26	No Health Consequences or Impact	A070101	Failure to Capture		
									A070102	High Capture Threshold		
									A071209	Extracardiac Stimulation		
									A1301	Failure to Read Input Signal		
					CT1685	Cardiac pacemakers	E2103	Device Overstimulation of Tissue	A070101	Failure to Capture		
									A070102	High Capture Threshold		

									A071209	Extracardiac Stimulation		
									A1301	Failure to Read Input Signal		
							F26	No Health Consequences or Impact	A070101	Failure to Capture		
									A070102	High Capture Threshold		
									A071209	Extracardiac Stimulation		
									A1301	Failure to Read Input Signal		
					CT233	Surgical	E2103	Device Overstimulation of Tissue	A070101	Failure to Capture		
									A070102	High Capture Threshold		
									A071209	Extracardiac Stimulation		
									A1301	Failure to Read Input Signal		
							F26	No Health Consequences or Impact	A070101	Failure to Capture		
									A070102	High Capture Threshold		
									A071209	Extracardiac Stimulation		
									A1301	Failure to Read Input Signal		
					CT2406	Implantable	E2103	Device Overstimulation of Tissue	A070101	Failure to Capture		
									A070102	High Capture Threshold		
									A071209	Extracardiac Stimulation		
									A1301	Failure to Read Input Signal		
							F26	No Health Consequences or Impact	A070101	Failure to Capture		
									A070102	High Capture Threshold		
									A071209	Extracardiac Stimulation		
									A1301	Failure to Read Input Signal		
					CT2407	Active implantable	E2103	Device Overstimulation of Tissue	A070101	Failure to Capture		
									A070102	High Capture Threshold		
									A071209	Extracardiac Stimulation		
									A1301	Failure to Read Input Signal		
							F26	No Health Consequences or Impact	A070101	Failure to Capture		
									A070102	High Capture Threshold		
									A071209	Extracardiac Stimulation		
									A1301	Failure to Read Input Signal		
					CT257	Implantable cardiac pacemakers	E2103	Device Overstimulation of Tissue	A070101	Failure to Capture		
									A070102	High Capture Threshold		
									A071209	Extracardiac Stimulation		
									A1301	Failure to Read Input Signal		
							F26	No Health Consequences or Impact	A070101	Failure to Capture		
									A070102	High Capture Threshold		
									A071209	Extracardiac Stimulation		
									A1301	Failure to Read Input Signal		
					CT2727	Active implant devices/ancillaries	E2103	Device Overstimulation of Tissue	A070101	Failure to Capture		
									A070102	High Capture Threshold		
									A071209	Extracardiac Stimulation		
									A1301	Failure to Read Input Signal		
							F26	No Health Consequences or Impact	A070101	Failure to Capture		
									A070102	High Capture Threshold		
									A071209	Extracardiac Stimulation		
									A1301	Failure to Read Input Signal		
					CT2750	Implantable cardiac electrical stimulators and associated devices	E2103	Device Overstimulation of Tissue	A070101	Failure to Capture		
									A070102	High Capture Threshold		
									A071209	Extracardiac Stimulation		
									A1301	Failure to Read Input Signal		
							F26	No Health Consequences or Impact	A070101	Failure to Capture		
									A070102	High Capture Threshold		
									A071209	Extracardiac Stimulation		
									A1301	Failure to Read Input Signal		
					CT2866	Implantable cardiac electrical stimulators	E2103	Device Overstimulation of Tissue	A070101	Failure to Capture		
									A070102	High Capture Threshold		
									A071209	Extracardiac Stimulation		
									A1301	Failure to Read Input Signal		
							F26	No Health Consequences or Impact	A070101	Failure to Capture		
									A070102	High Capture Threshold		
									A071209	Extracardiac Stimulation		
									A1301	Failure to Read Input Signal		
					CT2985	Active	E2103	Device Overstimulation of Tissue	A070101	Failure to Capture		
									A070102	High Capture Threshold		
									A071209	Extracardiac Stimulation		
									A1301	Failure to Read Input Signal		
							F26	No Health Consequences or Impact	A070101	Failure to Capture		
									A070102	High Capture Threshold		
									A071209	Extracardiac Stimulation		
									A1301	Failure to Read Input Signal		
					CT3001	Implantable metal	E2103	Device Overstimulation of Tissue	A070101	Failure to Capture		
									A070102	High Capture Threshold		
									A071209	Extracardiac Stimulation		
									A1301	Failure to Read Input Signal		
							F26	No Health Consequences or Impact	A070101	Failure to Capture		
									A070102	High Capture Threshold		
									A071209	Extracardiac Stimulation		
									A1301	Failure to Read Input Signal		
					CT3002	Implantable synthetic polymer	E2103	Device Overstimulation of Tissue	A070101	Failure to Capture		
									A070102	High Capture Threshold		
									A071209	Extracardiac Stimulation		
									A1301	Failure to Read Input Signal		
							F26	No Health Consequences or Impact	A070101	Failure to Capture		
									A070102	High Capture Threshold		
									A071209	Extracardiac Stimulation		
									A1301	Failure to Read Input Signal		
					CT3006	Implantable inorganic material	E2103	Device Overstimulation of Tissue	A070101	Failure to Capture		
									A070102	High Capture Threshold		

									A071209	Extracardiac Stimulation		
									A1301	Failure to Read Input Signal		
							F26	No Health Consequences or Impact	A070101	Failure to Capture		
									A070102	High Capture Threshold		
									A071209	Extracardiac Stimulation		
									A1301	Failure to Read Input Signal		
					CT314	Electrically powered	E2103	Device Overstimulation of Tissue	A070101	Failure to Capture		
									A070102	High Capture Threshold		
									A071209	Extracardiac Stimulation		
									A1301	Failure to Read Input Signal		
							F26	No Health Consequences or Impact	A070101	Failure to Capture		
									A070102	High Capture Threshold		
									A071209	Extracardiac Stimulation		
									A1301	Failure to Read Input Signal		
					CT321	Long-term surgical invasive	E2103	Device Overstimulation of Tissue	A070101	Failure to Capture		
									A070102	High Capture Threshold		
									A071209	Extracardiac Stimulation		
									A1301	Failure to Read Input Signal		
							F26	No Health Consequences or Impact	A070101	Failure to Capture		
									A070102	High Capture Threshold		
									A071209	Extracardiac Stimulation		
									A1301	Failure to Read Input Signal		
					CT449	Generators and associated devices	E2103	Device Overstimulation of Tissue	A070101	Failure to Capture		
									A070102	High Capture Threshold		
									A071209	Extracardiac Stimulation		
									A1301	Failure to Read Input Signal		
							F26	No Health Consequences or Impact	A070101	Failure to Capture		
									A070102	High Capture Threshold		
									A071209	Extracardiac Stimulation		
									A1301	Failure to Read Input Signal		
					CT981	Single use	E2103	Device Overstimulation of Tissue	A070101	Failure to Capture		
									A070102	High Capture Threshold		
									A071209	Extracardiac Stimulation		
									A1301	Failure to Read Input Signal		
							F26	No Health Consequences or Impact	A070101	Failure to Capture		
									A070102	High Capture Threshold		
									A071209	Extracardiac Stimulation		
									A1301	Failure to Read Input Signal		
					CT983	Surgical invasive	E2103	Device Overstimulation of Tissue	A070101	Failure to Capture		
									A070102	High Capture Threshold		
									A071209	Extracardiac Stimulation		
									A1301	Failure to Read Input Signal		
							F26	No Health Consequences or Impact	A070101	Failure to Capture		
									A070102	High Capture Threshold		
									A071209	Extracardiac Stimulation		
									A1301	Failure to Read Input Signal		
13573	2023		47265	Dual-chamber implantable pacemaker, rate-responsive	CT141	Electrodes/Leads and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A090805	Incorrect Measurement		
							F26	No Health Consequences or Impact	A090805	Incorrect Measurement		
					CT145	Cardiovascular devices	E2403	No Clinical Signs, Symptoms or Conditions	A090805	Incorrect Measurement		
							F26	No Health Consequences or Impact	A090805	Incorrect Measurement		
					CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A090805	Incorrect Measurement		
							F26	No Health Consequences or Impact	A090805	Incorrect Measurement		
					CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A090805	Incorrect Measurement		
							F26	No Health Consequences or Impact	A090805	Incorrect Measurement		
					CT233	Surgical	E2403	No Clinical Signs, Symptoms or Conditions	A090805	Incorrect Measurement		
							F26	No Health Consequences or Impact	A090805	Incorrect Measurement		
					CT2406	Implantable	E2403	No Clinical Signs, Symptoms or Conditions	A090805	Incorrect Measurement		
							F26	No Health Consequences or Impact	A090805	Incorrect Measurement		
					CT2407	Active implantable	E2403	No Clinical Signs, Symptoms or Conditions	A090805	Incorrect Measurement		
							F26	No Health Consequences or Impact	A090805	Incorrect Measurement		
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A090805	Incorrect Measurement		
							F26	No Health Consequences or Impact	A090805	Incorrect Measurement		
					CT2727	Active implant devices/ancillaries	E2403	No Clinical Signs, Symptoms or Conditions	A090805	Incorrect Measurement		
							F26	No Health Consequences or Impact	A090805	Incorrect Measurement		
					CT2750	Implantable cardiac electrical stimulators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A090805	Incorrect Measurement		
							F26	No Health Consequences or Impact	A090805	Incorrect Measurement		
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A090805	Incorrect Measurement		
							F26	No Health Consequences or Impact	A090805	Incorrect Measurement		
					CT2985	Active	E2403	No Clinical Signs, Symptoms or Conditions	A090805	Incorrect Measurement		
							F26	No Health Consequences or Impact	A090805	Incorrect Measurement		
					CT3001	Implantable metal	E2403	No Clinical Signs, Symptoms or Conditions	A090805	Incorrect Measurement		
							F26	No Health Consequences or Impact	A090805	Incorrect Measurement		
					CT3002	Implantable synthetic polymer	E2403	No Clinical Signs, Symptoms or Conditions	A090805	Incorrect Measurement		
							F26	No Health Consequences or Impact	A090805	Incorrect Measurement		
					CT3006	Implantable inorganic material	E2403	No Clinical Signs, Symptoms or Conditions	A090805	Incorrect Measurement		
							F26	No Health Consequences or Impact	A090805	Incorrect Measurement		
					CT314	Electrically powered	E2403	No Clinical Signs, Symptoms or Conditions	A090805	Incorrect Measurement		
							F26	No Health Consequences or Impact	A090805	Incorrect Measurement		
					CT321	Long-term surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A090805	Incorrect Measurement		
							F26	No Health Consequences or Impact	A090805	Incorrect Measurement		
					CT449	Generators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A090805	Incorrect Measurement		
							F26	No Health Consequences or Impact	A090805	Incorrect Measurement		
					CT981	Single use	E2403	No Clinical Signs, Symptoms or Conditions	A090805	Incorrect Measurement		
							F26	No Health Consequences or Impact	A090805	Incorrect Measurement		
					CT983	Surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A090805	Incorrect Measurement		
							F26	No Health Consequences or Impact	A090805	Incorrect Measurement		
13574	2023	All other reportable incidents	47265	Dual-chamber implantable pacemaker, rate-responsive	CT141	Electrodes/Leads and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A071102	Failure to Interrogate	11/07/2023	11/07/2023
							F1901	Additional Surgery	A071102	Failure to Interrogate	11/07/2023	11/07/2023

					CT145	Cardiovascular devices	E2403	No Clinical Signs, Symptoms or Conditions	A071102	Failure to Interrogate	11/07/2023	11/07/2023
							F1901	Additional Surgery	A071102	Failure to Interrogate	11/07/2023	11/07/2023
					CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A071102	Failure to Interrogate	11/07/2023	11/07/2023
							F1901	Additional Surgery	A071102	Failure to Interrogate	11/07/2023	11/07/2023
					CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A071102	Failure to Interrogate	11/07/2023	11/07/2023
							F1901	Additional Surgery	A071102	Failure to Interrogate	11/07/2023	11/07/2023
					CT233	Surgical	E2403	No Clinical Signs, Symptoms or Conditions	A071102	Failure to Interrogate	11/07/2023	11/07/2023
							F1901	Additional Surgery	A071102	Failure to Interrogate	11/07/2023	11/07/2023
					CT2406	Implantable	E2403	No Clinical Signs, Symptoms or Conditions	A071102	Failure to Interrogate	11/07/2023	11/07/2023
							F1901	Additional Surgery	A071102	Failure to Interrogate	11/07/2023	11/07/2023
					CT2407	Active implantable	E2403	No Clinical Signs, Symptoms or Conditions	A071102	Failure to Interrogate	11/07/2023	11/07/2023
							F1901	Additional Surgery	A071102	Failure to Interrogate	11/07/2023	11/07/2023
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A071102	Failure to Interrogate	11/07/2023	11/07/2023
							F1901	Additional Surgery	A071102	Failure to Interrogate	11/07/2023	11/07/2023
					CT2727	Active implant devices/ancillaries	E2403	No Clinical Signs, Symptoms or Conditions	A071102	Failure to Interrogate	11/07/2023	11/07/2023
							F1901	Additional Surgery	A071102	Failure to Interrogate	11/07/2023	11/07/2023
					CT2750	Implantable cardiac electrical stimulators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A071102	Failure to Interrogate	11/07/2023	11/07/2023
							F1901	Additional Surgery	A071102	Failure to Interrogate	11/07/2023	11/07/2023
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A071102	Failure to Interrogate	11/07/2023	11/07/2023
							F1901	Additional Surgery	A071102	Failure to Interrogate	11/07/2023	11/07/2023
					CT2985	Active	E2403	No Clinical Signs, Symptoms or Conditions	A071102	Failure to Interrogate	11/07/2023	11/07/2023
							F1901	Additional Surgery	A071102	Failure to Interrogate	11/07/2023	11/07/2023
					CT3001	Implantable metal	E2403	No Clinical Signs, Symptoms or Conditions	A071102	Failure to Interrogate	11/07/2023	11/07/2023
							F1901	Additional Surgery	A071102	Failure to Interrogate	11/07/2023	11/07/2023
					CT3002	Implantable synthetic polymer	E2403	No Clinical Signs, Symptoms or Conditions	A071102	Failure to Interrogate	11/07/2023	11/07/2023
							F1901	Additional Surgery	A071102	Failure to Interrogate	11/07/2023	11/07/2023
					CT3006	Implantable inorganic material	E2403	No Clinical Signs, Symptoms or Conditions	A071102	Failure to Interrogate	11/07/2023	11/07/2023
							F1901	Additional Surgery	A071102	Failure to Interrogate	11/07/2023	11/07/2023
					CT314	Electrically powered	E2403	No Clinical Signs, Symptoms or Conditions	A071102	Failure to Interrogate	11/07/2023	11/07/2023
							F1901	Additional Surgery	A071102	Failure to Interrogate	11/07/2023	11/07/2023
					CT321	Long-term surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A071102	Failure to Interrogate	11/07/2023	11/07/2023
							F1901	Additional Surgery	A071102	Failure to Interrogate	11/07/2023	11/07/2023
					CT449	Generators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A071102	Failure to Interrogate	11/07/2023	11/07/2023
							F1901	Additional Surgery	A071102	Failure to Interrogate	11/07/2023	11/07/2023
					CT981	Single use	E2403	No Clinical Signs, Symptoms or Conditions	A071102	Failure to Interrogate	11/07/2023	11/07/2023
							F1901	Additional Surgery	A071102	Failure to Interrogate	11/07/2023	11/07/2023
					CT983	Surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A071102	Failure to Interrogate	11/07/2023	11/07/2023
							F1901	Additional Surgery	A071102	Failure to Interrogate	11/07/2023	11/07/2023
13575	2023	All other reportable incidents	47263	Cardiac resynchronization therapy implantable pacemaker	CT1684	Cardiac pacemakers and associated devices	E0602	Cardiac Arrest	A070101	Failure to Capture	17/06/2023	17/06/2023
									A0908	Incorrect, Inadequate or Imprecise Result or Readings	17/06/2023	17/06/2023
							F08	Hospitalization or Prolonged Hospitalization	A070101	Failure to Capture	17/06/2023	17/06/2023
									A0908	Incorrect, Inadequate or Imprecise Result or Readings	17/06/2023	17/06/2023
							F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A070101	Failure to Capture	17/06/2023	17/06/2023
									A0908	Incorrect, Inadequate or Imprecise Result or Readings	17/06/2023	17/06/2023
							F1905	Device Revision or Replacement	A070101	Failure to Capture	17/06/2023	17/06/2023
									A0908	Incorrect, Inadequate or Imprecise Result or Readings	17/06/2023	17/06/2023
							F23	Unexpected Medical Intervention	A070101	Failure to Capture	17/06/2023	17/06/2023
									A0908	Incorrect, Inadequate or Imprecise Result or Readings	17/06/2023	17/06/2023
					CT1685	Cardiac pacemakers	E0602	Cardiac Arrest	A070101	Failure to Capture	17/06/2023	17/06/2023
									A0908	Incorrect, Inadequate or Imprecise Result or Readings	17/06/2023	17/06/2023
							F08	Hospitalization or Prolonged Hospitalization	A070101	Failure to Capture	17/06/2023	17/06/2023
									A0908	Incorrect, Inadequate or Imprecise Result or Readings	17/06/2023	17/06/2023
							F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A070101	Failure to Capture	17/06/2023	17/06/2023
									A0908	Incorrect, Inadequate or Imprecise Result or Readings	17/06/2023	17/06/2023
							F1905	Device Revision or Replacement	A070101	Failure to Capture	17/06/2023	17/06/2023
									A0908	Incorrect, Inadequate or Imprecise Result or Readings	17/06/2023	17/06/2023
							F23	Unexpected Medical Intervention	A070101	Failure to Capture	17/06/2023	17/06/2023
									A0908	Incorrect, Inadequate or Imprecise Result or Readings	17/06/2023	17/06/2023
					CT257	Implantable cardiac pacemakers	E0602	Cardiac Arrest	A070101	Failure to Capture	17/06/2023	17/06/2023
									A0908	Incorrect, Inadequate or Imprecise Result or Readings	17/06/2023	17/06/2023
							F08	Hospitalization or Prolonged Hospitalization	A070101	Failure to Capture	17/06/2023	17/06/2023
									A0908	Incorrect, Inadequate or Imprecise Result or Readings	17/06/2023	17/06/2023
							F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A070101	Failure to Capture	17/06/2023	17/06/2023
									A0908	Incorrect, Inadequate or Imprecise Result or Readings	17/06/2023	17/06/2023
							F1905	Device Revision or Replacement	A070101	Failure to Capture	17/06/2023	17/06/2023
									A0908	Incorrect, Inadequate or Imprecise Result or Readings	17/06/2023	17/06/2023
							F23	Unexpected Medical Intervention	A070101	Failure to Capture	17/06/2023	17/06/2023
									A0908	Incorrect, Inadequate or Imprecise Result or Readings	17/06/2023	17/06/2023
					CT2866	Implantable cardiac electrical stimulators	E0602	Cardiac Arrest	A070101	Failure to Capture	17/06/2023	17/06/2023
									A0908	Incorrect, Inadequate or Imprecise Result or Readings	17/06/2023	17/06/2023
							F08	Hospitalization or Prolonged Hospitalization	A070101	Failure to Capture	17/06/2023	17/06/2023
									A0908	Incorrect, Inadequate or Imprecise Result or Readings	17/06/2023	17/06/2023
							F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A070101	Failure to Capture	17/06/2023	17/06/2023
									A0908	Incorrect, Inadequate or Imprecise Result or Readings	17/06/2023	17/06/2023

							F1905	Device Revision or Replacement	A070101	Failure to Capture	17/06/2023	17/06/2023
									A0908	Incorrect, Inadequate or Imprecise Result or Readings	17/06/2023	17/06/2023
							F23	Unexpected Medical Intervention	A070101	Failure to Capture	17/06/2023	17/06/2023
									A0908	Incorrect, Inadequate or Imprecise Result or Readings	17/06/2023	17/06/2023
13576	2023	All other reportable incidents	47265	Dual-chamber implantable pacemaker, rate-responsive	CT141	Electrodes/Leads and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	28/06/2023	28/06/2023
									A0908	Incorrect, Inadequate or Imprecise Result or Readings	28/06/2023	28/06/2023
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	28/06/2023	28/06/2023
									A0908	Incorrect, Inadequate or Imprecise Result or Readings	28/06/2023	28/06/2023
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	28/06/2023	28/06/2023
									A0908	Incorrect, Inadequate or Imprecise Result or Readings	28/06/2023	28/06/2023
					CT145	Cardiovascular devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	28/06/2023	28/06/2023
									A0908	Incorrect, Inadequate or Imprecise Result or Readings	28/06/2023	28/06/2023
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	28/06/2023	28/06/2023
									A0908	Incorrect, Inadequate or Imprecise Result or Readings	28/06/2023	28/06/2023
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	28/06/2023	28/06/2023
									A0908	Incorrect, Inadequate or Imprecise Result or Readings	28/06/2023	28/06/2023
					CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	28/06/2023	28/06/2023
									A0908	Incorrect, Inadequate or Imprecise Result or Readings	28/06/2023	28/06/2023
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	28/06/2023	28/06/2023
									A0908	Incorrect, Inadequate or Imprecise Result or Readings	28/06/2023	28/06/2023
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	28/06/2023	28/06/2023
									A0908	Incorrect, Inadequate or Imprecise Result or Readings	28/06/2023	28/06/2023
					CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	28/06/2023	28/06/2023
									A0908	Incorrect, Inadequate or Imprecise Result or Readings	28/06/2023	28/06/2023
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	28/06/2023	28/06/2023
									A0908	Incorrect, Inadequate or Imprecise Result or Readings	28/06/2023	28/06/2023
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	28/06/2023	28/06/2023
									A0908	Incorrect, Inadequate or Imprecise Result or Readings	28/06/2023	28/06/2023
					CT233	Surgical	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	28/06/2023	28/06/2023
									A0908	Incorrect, Inadequate or Imprecise Result or Readings	28/06/2023	28/06/2023
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	28/06/2023	28/06/2023
									A0908	Incorrect, Inadequate or Imprecise Result or Readings	28/06/2023	28/06/2023
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	28/06/2023	28/06/2023
									A0908	Incorrect, Inadequate or Imprecise Result or Readings	28/06/2023	28/06/2023
					CT2406	Implantable	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	28/06/2023	28/06/2023
									A0908	Incorrect, Inadequate or Imprecise Result or Readings	28/06/2023	28/06/2023
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	28/06/2023	28/06/2023
									A0908	Incorrect, Inadequate or Imprecise Result or Readings	28/06/2023	28/06/2023
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	28/06/2023	28/06/2023
									A0908	Incorrect, Inadequate or Imprecise Result or Readings	28/06/2023	28/06/2023
					CT2407	Active implantable	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	28/06/2023	28/06/2023
									A0908	Incorrect, Inadequate or Imprecise Result or Readings	28/06/2023	28/06/2023
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	28/06/2023	28/06/2023
									A0908	Incorrect, Inadequate or Imprecise Result or Readings	28/06/2023	28/06/2023
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	28/06/2023	28/06/2023
									A0908	Incorrect, Inadequate or Imprecise Result or Readings	28/06/2023	28/06/2023
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	28/06/2023	28/06/2023
									A0908	Incorrect, Inadequate or Imprecise Result or Readings	28/06/2023	28/06/2023
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	28/06/2023	28/06/2023
									A0908	Incorrect, Inadequate or Imprecise Result or Readings	28/06/2023	28/06/2023
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	28/06/2023	28/06/2023
									A0908	Incorrect, Inadequate or Imprecise Result or Readings	28/06/2023	28/06/2023
					CT2727	Active implant devices/ancillaries	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	28/06/2023	28/06/2023
									A0908	Incorrect, Inadequate or Imprecise Result or Readings	28/06/2023	28/06/2023
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	28/06/2023	28/06/2023
									A0908	Incorrect, Inadequate or Imprecise Result or Readings	28/06/2023	28/06/2023
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	28/06/2023	28/06/2023
									A0908	Incorrect, Inadequate or Imprecise Result or Readings	28/06/2023	28/06/2023
					CT2750	Implantable cardiac electrical stimulators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	28/06/2023	28/06/2023
									A0908	Incorrect, Inadequate or Imprecise Result or Readings	28/06/2023	28/06/2023
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	28/06/2023	28/06/2023
									A0908	Incorrect, Inadequate or Imprecise Result or Readings	28/06/2023	28/06/2023
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	28/06/2023	28/06/2023

									A0908	Incorrect, Inadequate or Imprecise Result or Readings	28/06/2023	28/06/2023
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	28/06/2023	28/06/2023
									A0908	Incorrect, Inadequate or Imprecise Result or Readings	28/06/2023	28/06/2023
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	28/06/2023	28/06/2023
									A0908	Incorrect, Inadequate or Imprecise Result or Readings	28/06/2023	28/06/2023
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	28/06/2023	28/06/2023
									A0908	Incorrect, Inadequate or Imprecise Result or Readings	28/06/2023	28/06/2023
					CT2985	Active	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	28/06/2023	28/06/2023
									A0908	Incorrect, Inadequate or Imprecise Result or Readings	28/06/2023	28/06/2023
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	28/06/2023	28/06/2023
									A0908	Incorrect, Inadequate or Imprecise Result or Readings	28/06/2023	28/06/2023
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	28/06/2023	28/06/2023
									A0908	Incorrect, Inadequate or Imprecise Result or Readings	28/06/2023	28/06/2023
					CT3001	Implantable metal	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	28/06/2023	28/06/2023
									A0908	Incorrect, Inadequate or Imprecise Result or Readings	28/06/2023	28/06/2023
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	28/06/2023	28/06/2023
									A0908	Incorrect, Inadequate or Imprecise Result or Readings	28/06/2023	28/06/2023
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	28/06/2023	28/06/2023
									A0908	Incorrect, Inadequate or Imprecise Result or Readings	28/06/2023	28/06/2023
					CT3002	Implantable synthetic polymer	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	28/06/2023	28/06/2023
									A0908	Incorrect, Inadequate or Imprecise Result or Readings	28/06/2023	28/06/2023
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	28/06/2023	28/06/2023
									A0908	Incorrect, Inadequate or Imprecise Result or Readings	28/06/2023	28/06/2023
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	28/06/2023	28/06/2023
									A0908	Incorrect, Inadequate or Imprecise Result or Readings	28/06/2023	28/06/2023
					CT3006	Implantable inorganic material	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	28/06/2023	28/06/2023
									A0908	Incorrect, Inadequate or Imprecise Result or Readings	28/06/2023	28/06/2023
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	28/06/2023	28/06/2023
									A0908	Incorrect, Inadequate or Imprecise Result or Readings	28/06/2023	28/06/2023
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	28/06/2023	28/06/2023
									A0908	Incorrect, Inadequate or Imprecise Result or Readings	28/06/2023	28/06/2023
					CT314	Electrically powered	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	28/06/2023	28/06/2023
									A0908	Incorrect, Inadequate or Imprecise Result or Readings	28/06/2023	28/06/2023
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	28/06/2023	28/06/2023
									A0908	Incorrect, Inadequate or Imprecise Result or Readings	28/06/2023	28/06/2023
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	28/06/2023	28/06/2023
									A0908	Incorrect, Inadequate or Imprecise Result or Readings	28/06/2023	28/06/2023
					CT321	Long-term surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	28/06/2023	28/06/2023
									A0908	Incorrect, Inadequate or Imprecise Result or Readings	28/06/2023	28/06/2023
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	28/06/2023	28/06/2023
									A0908	Incorrect, Inadequate or Imprecise Result or Readings	28/06/2023	28/06/2023
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	28/06/2023	28/06/2023
									A0908	Incorrect, Inadequate or Imprecise Result or Readings	28/06/2023	28/06/2023
					CT449	Generators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	28/06/2023	28/06/2023
									A0908	Incorrect, Inadequate or Imprecise Result or Readings	28/06/2023	28/06/2023
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	28/06/2023	28/06/2023
									A0908	Incorrect, Inadequate or Imprecise Result or Readings	28/06/2023	28/06/2023
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	28/06/2023	28/06/2023
									A0908	Incorrect, Inadequate or Imprecise Result or Readings	28/06/2023	28/06/2023
					CT981	Single use	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	28/06/2023	28/06/2023
									A0908	Incorrect, Inadequate or Imprecise Result or Readings	28/06/2023	28/06/2023
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	28/06/2023	28/06/2023
									A0908	Incorrect, Inadequate or Imprecise Result or Readings	28/06/2023	28/06/2023
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	28/06/2023	28/06/2023
									A0908	Incorrect, Inadequate or Imprecise Result or Readings	28/06/2023	28/06/2023
					CT983	Surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	28/06/2023	28/06/2023
									A0908	Incorrect, Inadequate or Imprecise Result or Readings	28/06/2023	28/06/2023
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	28/06/2023	28/06/2023
									A0908	Incorrect, Inadequate or Imprecise Result or Readings	28/06/2023	28/06/2023
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	28/06/2023	28/06/2023
									A0908	Incorrect, Inadequate or Imprecise Result or Readings	28/06/2023	28/06/2023
13577	2023	All other reportable incidents	36241	Pacemaker/defibrillator lead	CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070910	Undersensing		
									A0712	Pacing Problem		
									A0909	Unexpected Therapeutic Results		

							F10	Inadequate/inappropriate Treatment or Diagnostic Exposure	A070910	Undersensing			
									A0712	Pacing Problem			
									A0909	Unexpected Therapeutic Results			
13578	2023		47265	Dual-chamber implantable pacemaker, rate-responsive	CT141	Electrodes/Leads and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem			
							F1901	Additional Surgery	A24	No Apparent Device Problem			
					CT145	Cardiovascular devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem			
							F1901	Additional Surgery	A24	No Apparent Device Problem			
					CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem			
							F1901	Additional Surgery	A24	No Apparent Device Problem			
					CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem			
							F1901	Additional Surgery	A24	No Apparent Device Problem			
					CT233	Surgical	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem			
							F1901	Additional Surgery	A24	No Apparent Device Problem			
					CT2406	Implantable	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem			
							F1901	Additional Surgery	A24	No Apparent Device Problem			
					CT2407	Active implantable	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem			
							F1901	Additional Surgery	A24	No Apparent Device Problem			
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem			
							F1901	Additional Surgery	A24	No Apparent Device Problem			
					CT2727	Active implant devices/ancillaries	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem			
							F1901	Additional Surgery	A24	No Apparent Device Problem			
					CT2750	Implantable cardiac electrical stimulators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem			
							F1901	Additional Surgery	A24	No Apparent Device Problem			
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem			
							F1901	Additional Surgery	A24	No Apparent Device Problem			
					CT2985	Active	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem			
							F1901	Additional Surgery	A24	No Apparent Device Problem			
					CT3001	Implantable metal	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem			
							F1901	Additional Surgery	A24	No Apparent Device Problem			
					CT3002	Implantable synthetic polymer	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem			
							F1901	Additional Surgery	A24	No Apparent Device Problem			
					CT3006	Implantable inorganic material	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem			
							F1901	Additional Surgery	A24	No Apparent Device Problem			
					CT314	Electrically powered	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem			
							F1901	Additional Surgery	A24	No Apparent Device Problem			
					CT321	Long-term surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem			
							F1901	Additional Surgery	A24	No Apparent Device Problem			
					CT449	Generators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem			
							F1901	Additional Surgery	A24	No Apparent Device Problem			
					CT981	Single use	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem			
							F1901	Additional Surgery	A24	No Apparent Device Problem			
					CT983	Surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem			
							F1901	Additional Surgery	A24	No Apparent Device Problem			
13579	2023	All other reportable incidents	47265	Dual-chamber implantable pacemaker, rate-responsive	CT141	Electrodes/Leads and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A071102	Failure to Interrogate			
									A13	Communication or Transmission Problem			
							F08	Hospitalization or Prolonged Hospitalization	A071102	Failure to Interrogate			
									A13	Communication or Transmission Problem			
							F12	Serious Injury/ Illness/ Impairment	A071102	Failure to Interrogate			
									A13	Communication or Transmission Problem			
							F1905	Device Revision or Replacement	A071102	Failure to Interrogate			
									A13	Communication or Transmission Problem			
							F23	Unexpected Medical Intervention	A071102	Failure to Interrogate			
									A13	Communication or Transmission Problem			
					CT145	Cardiovascular devices	E2403	No Clinical Signs, Symptoms or Conditions	A071102	Failure to Interrogate			
									A13	Communication or Transmission Problem			
							F08	Hospitalization or Prolonged Hospitalization	A071102	Failure to Interrogate			
									A13	Communication or Transmission Problem			
							F12	Serious Injury/ Illness/ Impairment	A071102	Failure to Interrogate			
									A13	Communication or Transmission Problem			
							F1905	Device Revision or Replacement	A071102	Failure to Interrogate			
									A13	Communication or Transmission Problem			
							F23	Unexpected Medical Intervention	A071102	Failure to Interrogate			
									A13	Communication or Transmission Problem			
					CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A071102	Failure to Interrogate			
									A13	Communication or Transmission Problem			
							F08	Hospitalization or Prolonged Hospitalization	A071102	Failure to Interrogate			
									A13	Communication or Transmission Problem			
							F12	Serious Injury/ Illness/ Impairment	A071102	Failure to Interrogate			
									A13	Communication or Transmission Problem			
							F1905	Device Revision or Replacement	A071102	Failure to Interrogate			
									A13	Communication or Transmission Problem			
							F23	Unexpected Medical Intervention	A071102	Failure to Interrogate			
									A13	Communication or Transmission Problem			
					CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A071102	Failure to Interrogate			
									A13	Communication or Transmission Problem			
							F08	Hospitalization or Prolonged Hospitalization	A071102	Failure to Interrogate			
									A13	Communication or Transmission Problem			
							F12	Serious Injury/ Illness/ Impairment	A071102	Failure to Interrogate			
									A13	Communication or Transmission Problem			
							F1905	Device Revision or Replacement	A071102	Failure to Interrogate			
									A13	Communication or Transmission Problem			
							F23	Unexpected Medical Intervention	A071102	Failure to Interrogate			
									A13	Communication or Transmission Problem			
					CT233	Surgical	E2403	No Clinical Signs, Symptoms or Conditions	A071102	Failure to Interrogate			
									A13	Communication or Transmission Problem			
							F08	Hospitalization or Prolonged Hospitalization	A071102	Failure to Interrogate			
									A13	Communication or Transmission Problem			
							F12	Serious Injury/ Illness/ Impairment	A071102	Failure to Interrogate			

									A13	Communication or Transmission Problem		
							F1905	Device Revision or Replacement	A071102	Failure to Interrogate		
									A13	Communication or Transmission Problem		
							F23	Unexpected Medical Intervention	A071102	Failure to Interrogate		
									A13	Communication or Transmission Problem		
					CT2406	Implantable	E2403	No Clinical Signs, Symptoms or Conditions	A071102	Failure to Interrogate		
									A13	Communication or Transmission Problem		
							F08	Hospitalization or Prolonged Hospitalization	A071102	Failure to Interrogate		
									A13	Communication or Transmission Problem		
							F12	Serious Injury/ Illness/ Impairment	A071102	Failure to Interrogate		
									A13	Communication or Transmission Problem		
							F1905	Device Revision or Replacement	A071102	Failure to Interrogate		
									A13	Communication or Transmission Problem		
							F23	Unexpected Medical Intervention	A071102	Failure to Interrogate		
									A13	Communication or Transmission Problem		
					CT2407	Active implantable	E2403	No Clinical Signs, Symptoms or Conditions	A071102	Failure to Interrogate		
									A13	Communication or Transmission Problem		
							F08	Hospitalization or Prolonged Hospitalization	A071102	Failure to Interrogate		
									A13	Communication or Transmission Problem		
							F12	Serious Injury/ Illness/ Impairment	A071102	Failure to Interrogate		
									A13	Communication or Transmission Problem		
							F1905	Device Revision or Replacement	A071102	Failure to Interrogate		
									A13	Communication or Transmission Problem		
							F23	Unexpected Medical Intervention	A071102	Failure to Interrogate		
									A13	Communication or Transmission Problem		
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A071102	Failure to Interrogate		
									A13	Communication or Transmission Problem		
							F08	Hospitalization or Prolonged Hospitalization	A071102	Failure to Interrogate		
									A13	Communication or Transmission Problem		
							F12	Serious Injury/ Illness/ Impairment	A071102	Failure to Interrogate		
									A13	Communication or Transmission Problem		
							F1905	Device Revision or Replacement	A071102	Failure to Interrogate		
									A13	Communication or Transmission Problem		
							F23	Unexpected Medical Intervention	A071102	Failure to Interrogate		
									A13	Communication or Transmission Problem		
					CT2727	Active implant devices/ancillaries	E2403	No Clinical Signs, Symptoms or Conditions	A071102	Failure to Interrogate		
									A13	Communication or Transmission Problem		
							F08	Hospitalization or Prolonged Hospitalization	A071102	Failure to Interrogate		
									A13	Communication or Transmission Problem		
							F12	Serious Injury/ Illness/ Impairment	A071102	Failure to Interrogate		
									A13	Communication or Transmission Problem		
							F1905	Device Revision or Replacement	A071102	Failure to Interrogate		
									A13	Communication or Transmission Problem		
							F23	Unexpected Medical Intervention	A071102	Failure to Interrogate		
									A13	Communication or Transmission Problem		
					CT2750	Implantable cardiac electrical stimulators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A071102	Failure to Interrogate		
									A13	Communication or Transmission Problem		
							F08	Hospitalization or Prolonged Hospitalization	A071102	Failure to Interrogate		
									A13	Communication or Transmission Problem		
							F12	Serious Injury/ Illness/ Impairment	A071102	Failure to Interrogate		
									A13	Communication or Transmission Problem		
							F1905	Device Revision or Replacement	A071102	Failure to Interrogate		
									A13	Communication or Transmission Problem		
							F23	Unexpected Medical Intervention	A071102	Failure to Interrogate		
									A13	Communication or Transmission Problem		
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A071102	Failure to Interrogate		
									A13	Communication or Transmission Problem		
							F08	Hospitalization or Prolonged Hospitalization	A071102	Failure to Interrogate		
									A13	Communication or Transmission Problem		
							F12	Serious Injury/ Illness/ Impairment	A071102	Failure to Interrogate		
									A13	Communication or Transmission Problem		
							F1905	Device Revision or Replacement	A071102	Failure to Interrogate		
									A13	Communication or Transmission Problem		
							F23	Unexpected Medical Intervention	A071102	Failure to Interrogate		
									A13	Communication or Transmission Problem		
					CT2985	Active	E2403	No Clinical Signs, Symptoms or Conditions	A071102	Failure to Interrogate		
									A13	Communication or Transmission Problem		
							F08	Hospitalization or Prolonged Hospitalization	A071102	Failure to Interrogate		
									A13	Communication or Transmission Problem		
							F12	Serious Injury/ Illness/ Impairment	A071102	Failure to Interrogate		
									A13	Communication or Transmission Problem		
							F1905	Device Revision or Replacement	A071102	Failure to Interrogate		
									A13	Communication or Transmission Problem		
							F23	Unexpected Medical Intervention	A071102	Failure to Interrogate		
									A13	Communication or Transmission Problem		
					CT3001	Implantable metal	E2403	No Clinical Signs, Symptoms or Conditions	A071102	Failure to Interrogate		
									A13	Communication or Transmission Problem		
							F08	Hospitalization or Prolonged Hospitalization	A071102	Failure to Interrogate		
									A13	Communication or Transmission Problem		
							F12	Serious Injury/ Illness/ Impairment	A071102	Failure to Interrogate		
									A13	Communication or Transmission Problem		
							F1905	Device Revision or Replacement	A071102	Failure to Interrogate		
									A13	Communication or Transmission Problem		
							F23	Unexpected Medical Intervention	A071102	Failure to Interrogate		
									A13	Communication or Transmission Problem		
					CT3002	Implantable synthetic polymer	E2403	No Clinical Signs, Symptoms or Conditions	A071102	Failure to Interrogate		
									A13	Communication or Transmission Problem		
							F08	Hospitalization or Prolonged Hospitalization	A071102	Failure to Interrogate		

									A13	Communication or Transmission Problem		
							F12	Serious Injury/ Illness/ Impairment	A071102	Failure to Interrogate		
									A13	Communication or Transmission Problem		
							F1905	Device Revision or Replacement	A071102	Failure to Interrogate		
									A13	Communication or Transmission Problem		
							F23	Unexpected Medical Intervention	A071102	Failure to Interrogate		
									A13	Communication or Transmission Problem		
					CT3006	Implantable inorganic material	E2403	No Clinical Signs, Symptoms or Conditions	A071102	Failure to Interrogate		
									A13	Communication or Transmission Problem		
							F08	Hospitalization or Prolonged Hospitalization	A071102	Failure to Interrogate		
									A13	Communication or Transmission Problem		
							F12	Serious Injury/ Illness/ Impairment	A071102	Failure to Interrogate		
									A13	Communication or Transmission Problem		
							F1905	Device Revision or Replacement	A071102	Failure to Interrogate		
									A13	Communication or Transmission Problem		
							F23	Unexpected Medical Intervention	A071102	Failure to Interrogate		
									A13	Communication or Transmission Problem		
					CT314	Electrically powered	E2403	No Clinical Signs, Symptoms or Conditions	A071102	Failure to Interrogate		
									A13	Communication or Transmission Problem		
							F08	Hospitalization or Prolonged Hospitalization	A071102	Failure to Interrogate		
									A13	Communication or Transmission Problem		
							F12	Serious Injury/ Illness/ Impairment	A071102	Failure to Interrogate		
									A13	Communication or Transmission Problem		
							F1905	Device Revision or Replacement	A071102	Failure to Interrogate		
									A13	Communication or Transmission Problem		
							F23	Unexpected Medical Intervention	A071102	Failure to Interrogate		
									A13	Communication or Transmission Problem		
					CT321	Long-term surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A071102	Failure to Interrogate		
									A13	Communication or Transmission Problem		
							F08	Hospitalization or Prolonged Hospitalization	A071102	Failure to Interrogate		
									A13	Communication or Transmission Problem		
							F12	Serious Injury/ Illness/ Impairment	A071102	Failure to Interrogate		
									A13	Communication or Transmission Problem		
							F1905	Device Revision or Replacement	A071102	Failure to Interrogate		
									A13	Communication or Transmission Problem		
							F23	Unexpected Medical Intervention	A071102	Failure to Interrogate		
									A13	Communication or Transmission Problem		
					CT449	Generators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A071102	Failure to Interrogate		
									A13	Communication or Transmission Problem		
							F08	Hospitalization or Prolonged Hospitalization	A071102	Failure to Interrogate		
									A13	Communication or Transmission Problem		
							F12	Serious Injury/ Illness/ Impairment	A071102	Failure to Interrogate		
									A13	Communication or Transmission Problem		
							F1905	Device Revision or Replacement	A071102	Failure to Interrogate		
									A13	Communication or Transmission Problem		
							F23	Unexpected Medical Intervention	A071102	Failure to Interrogate		
									A13	Communication or Transmission Problem		
					CT981	Single use	E2403	No Clinical Signs, Symptoms or Conditions	A071102	Failure to Interrogate		
									A13	Communication or Transmission Problem		
							F08	Hospitalization or Prolonged Hospitalization	A071102	Failure to Interrogate		
									A13	Communication or Transmission Problem		
							F12	Serious Injury/ Illness/ Impairment	A071102	Failure to Interrogate		
									A13	Communication or Transmission Problem		
							F1905	Device Revision or Replacement	A071102	Failure to Interrogate		
									A13	Communication or Transmission Problem		
							F23	Unexpected Medical Intervention	A071102	Failure to Interrogate		
									A13	Communication or Transmission Problem		
					CT983	Surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A071102	Failure to Interrogate		
									A13	Communication or Transmission Problem		
							F08	Hospitalization or Prolonged Hospitalization	A071102	Failure to Interrogate		
									A13	Communication or Transmission Problem		
							F12	Serious Injury/ Illness/ Impairment	A071102	Failure to Interrogate		
									A13	Communication or Transmission Problem		
							F1905	Device Revision or Replacement	A071102	Failure to Interrogate		
									A13	Communication or Transmission Problem		
							F23	Unexpected Medical Intervention	A071102	Failure to Interrogate		
									A13	Communication or Transmission Problem		
13580	2023	All other reportable incidents	47265	Dual-chamber implantable pacemaker, rate-responsive	CT141	Electrodes/Leads and associated devices	E1906	Unspecified Infection	A24	No Apparent Device Problem	14/07/2023	14/07/2023
							E2340	Wound Dehiscence	A24	No Apparent Device Problem	14/07/2023	14/07/2023
							F08	Hospitalization or Prolonged Hospitalization	A24	No Apparent Device Problem	14/07/2023	14/07/2023
							F12	Serious Injury/ Illness/ Impairment	A24	No Apparent Device Problem	14/07/2023	14/07/2023
							F1903	Device Explantation	A24	No Apparent Device Problem	14/07/2023	14/07/2023
							F2303	Medication Required	A24	No Apparent Device Problem	14/07/2023	14/07/2023
					CT145	Cardiovascular devices	E1906	Unspecified Infection	A24	No Apparent Device Problem	14/07/2023	14/07/2023
							E2340	Wound Dehiscence	A24	No Apparent Device Problem	14/07/2023	14/07/2023
							F08	Hospitalization or Prolonged Hospitalization	A24	No Apparent Device Problem	14/07/2023	14/07/2023
							F12	Serious Injury/ Illness/ Impairment	A24	No Apparent Device Problem	14/07/2023	14/07/2023
							F1903	Device Explantation	A24	No Apparent Device Problem	14/07/2023	14/07/2023
							F2303	Medication Required	A24	No Apparent Device Problem	14/07/2023	14/07/2023
					CT1684	Cardiac pacemakers and associated devices	E1906	Unspecified Infection	A24	No Apparent Device Problem	14/07/2023	14/07/2023
							E2340	Wound Dehiscence	A24	No Apparent Device Problem	14/07/2023	14/07/2023
							F08	Hospitalization or Prolonged Hospitalization	A24	No Apparent Device Problem	14/07/2023	14/07/2023
							F12	Serious Injury/ Illness/ Impairment	A24	No Apparent Device Problem	14/07/2023	14/07/2023
							F1903	Device Explantation	A24	No Apparent Device Problem	14/07/2023	14/07/2023
							F2303	Medication Required	A24	No Apparent Device Problem	14/07/2023	14/07/2023
					CT1685	Cardiac pacemakers	E1906	Unspecified Infection	A24	No Apparent Device Problem	14/07/2023	14/07/2023
							E2340	Wound Dehiscence	A24	No Apparent Device Problem	14/07/2023	14/07/2023
							F08	Hospitalization or Prolonged Hospitalization	A24	No Apparent Device Problem	14/07/2023	14/07/2023

							F12	Serious Injury/ Illness/ Impairment	A24	No Apparent Device Problem	14/07/2023	14/07/2023
							F1903	Device Explantation	A24	No Apparent Device Problem	14/07/2023	14/07/2023
							F2303	Medication Required	A24	No Apparent Device Problem	14/07/2023	14/07/2023
						CT233	Surgical	E1906	Unspecified Infection	A24	No Apparent Device Problem	14/07/2023	14/07/2023
								E2340	Wound Dehiscence	A24	No Apparent Device Problem	14/07/2023	14/07/2023
								F08	Hospitalization or Prolonged Hospitalization	A24	No Apparent Device Problem	14/07/2023	14/07/2023
								F12	Serious Injury/ Illness/ Impairment	A24	No Apparent Device Problem	14/07/2023	14/07/2023
								F1903	Device Explantation	A24	No Apparent Device Problem	14/07/2023	14/07/2023
								F2303	Medication Required	A24	No Apparent Device Problem	14/07/2023	14/07/2023
						CT2406	Implantable	E1906	Unspecified Infection	A24	No Apparent Device Problem	14/07/2023	14/07/2023
								E2340	Wound Dehiscence	A24	No Apparent Device Problem	14/07/2023	14/07/2023
								F08	Hospitalization or Prolonged Hospitalization	A24	No Apparent Device Problem	14/07/2023	14/07/2023
								F12	Serious Injury/ Illness/ Impairment	A24	No Apparent Device Problem	14/07/2023	14/07/2023
								F1903	Device Explantation	A24	No Apparent Device Problem	14/07/2023	14/07/2023
								F2303	Medication Required	A24	No Apparent Device Problem	14/07/2023	14/07/2023
						CT2407	Active implantable	E1906	Unspecified Infection	A24	No Apparent Device Problem	14/07/2023	14/07/2023
								E2340	Wound Dehiscence	A24	No Apparent Device Problem	14/07/2023	14/07/2023
								F08	Hospitalization or Prolonged Hospitalization	A24	No Apparent Device Problem	14/07/2023	14/07/2023
								F12	Serious Injury/ Illness/ Impairment	A24	No Apparent Device Problem	14/07/2023	14/07/2023
								F1903	Device Explantation	A24	No Apparent Device Problem	14/07/2023	14/07/2023
								F2303	Medication Required	A24	No Apparent Device Problem	14/07/2023	14/07/2023
						CT257	Implantable cardiac pacemakers	E1906	Unspecified Infection	A24	No Apparent Device Problem	14/07/2023	14/07/2023
								E2340	Wound Dehiscence	A24	No Apparent Device Problem	14/07/2023	14/07/2023
								F08	Hospitalization or Prolonged Hospitalization	A24	No Apparent Device Problem	14/07/2023	14/07/2023
								F12	Serious Injury/ Illness/ Impairment	A24	No Apparent Device Problem	14/07/2023	14/07/2023
								F1903	Device Explantation	A24	No Apparent Device Problem	14/07/2023	14/07/2023
								F2303	Medication Required	A24	No Apparent Device Problem	14/07/2023	14/07/2023
						CT2727	Active implant devices/ancillaries	E1906	Unspecified Infection	A24	No Apparent Device Problem	14/07/2023	14/07/2023
								E2340	Wound Dehiscence	A24	No Apparent Device Problem	14/07/2023	14/07/2023
								F08	Hospitalization or Prolonged Hospitalization	A24	No Apparent Device Problem	14/07/2023	14/07/2023
								F12	Serious Injury/ Illness/ Impairment	A24	No Apparent Device Problem	14/07/2023	14/07/2023
								F1903	Device Explantation	A24	No Apparent Device Problem	14/07/2023	14/07/2023
								F2303	Medication Required	A24	No Apparent Device Problem	14/07/2023	14/07/2023
						CT2750	Implantable cardiac electrical stimulators and associated devices	E1906	Unspecified Infection	A24	No Apparent Device Problem	14/07/2023	14/07/2023
								E2340	Wound Dehiscence	A24	No Apparent Device Problem	14/07/2023	14/07/2023
								F08	Hospitalization or Prolonged Hospitalization	A24	No Apparent Device Problem	14/07/2023	14/07/2023
								F12	Serious Injury/ Illness/ Impairment	A24	No Apparent Device Problem	14/07/2023	14/07/2023
								F1903	Device Explantation	A24	No Apparent Device Problem	14/07/2023	14/07/2023
								F2303	Medication Required	A24	No Apparent Device Problem	14/07/2023	14/07/2023
						CT2866	Implantable cardiac electrical stimulators	E1906	Unspecified Infection	A24	No Apparent Device Problem	14/07/2023	14/07/2023
								E2340	Wound Dehiscence	A24	No Apparent Device Problem	14/07/2023	14/07/2023
								F08	Hospitalization or Prolonged Hospitalization	A24	No Apparent Device Problem	14/07/2023	14/07/2023
								F12	Serious Injury/ Illness/ Impairment	A24	No Apparent Device Problem	14/07/2023	14/07/2023
								F1903	Device Explantation	A24	No Apparent Device Problem	14/07/2023	14/07/2023
								F2303	Medication Required	A24	No Apparent Device Problem	14/07/2023	14/07/2023
						CT2985	Active	E1906	Unspecified Infection	A24	No Apparent Device Problem	14/07/2023	14/07/2023
								E2340	Wound Dehiscence	A24	No Apparent Device Problem	14/07/2023	14/07/2023
								F08	Hospitalization or Prolonged Hospitalization	A24	No Apparent Device Problem	14/07/2023	14/07/2023
								F12	Serious Injury/ Illness/ Impairment	A24	No Apparent Device Problem	14/07/2023	14/07/2023
								F1903	Device Explantation	A24	No Apparent Device Problem	14/07/2023	14/07/2023
								F2303	Medication Required	A24	No Apparent Device Problem	14/07/2023	14/07/2023
						CT3001	Implantable metal	E1906	Unspecified Infection	A24	No Apparent Device Problem	14/07/2023	14/07/2023
								E2340	Wound Dehiscence	A24	No Apparent Device Problem	14/07/2023	14/07/2023
								F08	Hospitalization or Prolonged Hospitalization	A24	No Apparent Device Problem	14/07/2023	14/07/2023
								F12	Serious Injury/ Illness/ Impairment	A24	No Apparent Device Problem	14/07/2023	14/07/2023
								F1903	Device Explantation	A24	No Apparent Device Problem	14/07/2023	14/07/2023
								F2303	Medication Required	A24	No Apparent Device Problem	14/07/2023	14/07/2023
						CT3002	Implantable synthetic polymer	E1906	Unspecified Infection	A24	No Apparent Device Problem	14/07/2023	14/07/2023
								E2340	Wound Dehiscence	A24	No Apparent Device Problem	14/07/2023	14/07/2023
								F08	Hospitalization or Prolonged Hospitalization	A24	No Apparent Device Problem	14/07/2023	14/07/2023
								F12	Serious Injury/ Illness/ Impairment	A24	No Apparent Device Problem	14/07/2023	14/07/2023
								F1903	Device Explantation	A24	No Apparent Device Problem	14/07/2023	14/07/2023
								F2303	Medication Required	A24	No Apparent Device Problem	14/07/2023	14/07/2023
						CT3006	Implantable inorganic material	E1906	Unspecified Infection	A24	No Apparent Device Problem	14/07/2023	14/07/2023
								E2340	Wound Dehiscence	A24	No Apparent Device Problem	14/07/2023	14/07/2023
								F08	Hospitalization or Prolonged Hospitalization	A24	No Apparent Device Problem	14/07/2023	14/07/2023
								F12	Serious Injury/ Illness/ Impairment	A24	No Apparent Device Problem	14/07/2023	14/07/2023
								F1903	Device Explantation	A24	No Apparent Device Problem	14/07/2023	14/07/2023
								F2303	Medication Required	A24	No Apparent Device Problem	14/07/2023	14/07/2023
						CT314	Electrically powered	E1906	Unspecified Infection	A24	No Apparent Device Problem	14/07/2023	14/07/2023
								E2340	Wound Dehiscence	A24	No Apparent Device Problem	14/07/2023	14/07/2023
								F08	Hospitalization or Prolonged Hospitalization	A24	No Apparent Device Problem	14/07/2023	14/07/2023
								F12	Serious Injury/ Illness/ Impairment	A24	No Apparent Device Problem	14/07/2023	14/07/2023
								F1903	Device Explantation	A24	No Apparent Device Problem	14/07/2023	14/07/2023
								F2303	Medication Required	A24	No Apparent Device Problem	14/07/2023	14/07/2023
						CT321	Long-term surgical invasive	E1906	Unspecified Infection	A24	No Apparent Device Problem	14/07/2023	14/07/2023
								E2340	Wound Dehiscence	A24	No Apparent Device Problem	14/07/2023	14/07/2023
								F08	Hospitalization or Prolonged Hospitalization	A24	No Apparent Device Problem	14/07/2023	14/07/2023
								F12	Serious Injury/ Illness/ Impairment	A24	No Apparent Device Problem	14/07/2023	14/07/2023
								F1903	Device Explantation	A24	No Apparent Device Problem	14/07/2023	14/07/2023
								F2303	Medication Required	A24	No Apparent Device Problem	14/07/2023	14/07/2023
						CT449	Generators and associated devices	E1906	Unspecified Infection	A24	No Apparent Device Problem	14/07/2023	14/07/2023
								E2340	Wound Dehiscence	A24	No Apparent Device Problem	14/07/2023	14/07/2023
								F08	Hospitalization or Prolonged Hospitalization	A24	No Apparent Device Problem	14/07/2023	14/07/2023
								F12	Serious Injury/ Illness/ Impairment	A24	No Apparent Device Problem	14/07/2023	14/07/2023
								F1903	Device Explantation	A24	No Apparent Device Problem	14/07/2023	14/07/2023
								F2303	Medication Required	A24	No Apparent Device Problem	14/07/2023	14/07/2023
						CT981	Single use	E1906	Unspecified Infection	A24	No Apparent Device Problem	14/07/2023	14/07/2023

							E2340	Wound Dehiscence	A24	No Apparent Device Problem	14/07/2023	14/07/2023
							F08	Hospitalization or Prolonged Hospitalization	A24	No Apparent Device Problem	14/07/2023	14/07/2023
							F12	Serious Injury/ Illness/ Impairment	A24	No Apparent Device Problem	14/07/2023	14/07/2023
							F1903	Device Explantation	A24	No Apparent Device Problem	14/07/2023	14/07/2023
							F2303	Medication Required	A24	No Apparent Device Problem	14/07/2023	14/07/2023
					CT983	Surgical invasive	E1906	Unspecified Infection	A24	No Apparent Device Problem	14/07/2023	14/07/2023
							E2340	Wound Dehiscence	A24	No Apparent Device Problem	14/07/2023	14/07/2023
							F08	Hospitalization or Prolonged Hospitalization	A24	No Apparent Device Problem	14/07/2023	14/07/2023
							F12	Serious Injury/ Illness/ Impairment	A24	No Apparent Device Problem	14/07/2023	14/07/2023
							F1903	Device Explantation	A24	No Apparent Device Problem	14/07/2023	14/07/2023
							F2303	Medication Required	A24	No Apparent Device Problem	14/07/2023	14/07/2023
13581	2023	All other reportable incidents	47265	Dual-chamber implantable pacemaker, rate-responsive	CT141	Electrodes/Leads and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem	19/04/2023	19/04/2023
									A070504	Premature Discharge of Battery	19/04/2023	19/04/2023
							F08	Hospitalization or Prolonged Hospitalization	A0705	Battery Problem	19/04/2023	19/04/2023
									A070504	Premature Discharge of Battery	19/04/2023	19/04/2023
							F1905	Device Revision or Replacement	A0705	Battery Problem	19/04/2023	19/04/2023
									A070504	Premature Discharge of Battery	19/04/2023	19/04/2023
					CT145	Cardiovascular devices	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem	19/04/2023	19/04/2023
									A070504	Premature Discharge of Battery	19/04/2023	19/04/2023
							F08	Hospitalization or Prolonged Hospitalization	A0705	Battery Problem	19/04/2023	19/04/2023
									A070504	Premature Discharge of Battery	19/04/2023	19/04/2023
							F1905	Device Revision or Replacement	A0705	Battery Problem	19/04/2023	19/04/2023
									A070504	Premature Discharge of Battery	19/04/2023	19/04/2023
					CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem	19/04/2023	19/04/2023
									A070504	Premature Discharge of Battery	19/04/2023	19/04/2023
							F08	Hospitalization or Prolonged Hospitalization	A0705	Battery Problem	19/04/2023	19/04/2023
									A070504	Premature Discharge of Battery	19/04/2023	19/04/2023
							F1905	Device Revision or Replacement	A0705	Battery Problem	19/04/2023	19/04/2023
									A070504	Premature Discharge of Battery	19/04/2023	19/04/2023
					CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem	19/04/2023	19/04/2023
									A070504	Premature Discharge of Battery	19/04/2023	19/04/2023
							F08	Hospitalization or Prolonged Hospitalization	A0705	Battery Problem	19/04/2023	19/04/2023
									A070504	Premature Discharge of Battery	19/04/2023	19/04/2023
							F1905	Device Revision or Replacement	A0705	Battery Problem	19/04/2023	19/04/2023
									A070504	Premature Discharge of Battery	19/04/2023	19/04/2023
					CT233	Surgical	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem	19/04/2023	19/04/2023
									A070504	Premature Discharge of Battery	19/04/2023	19/04/2023
							F08	Hospitalization or Prolonged Hospitalization	A0705	Battery Problem	19/04/2023	19/04/2023
									A070504	Premature Discharge of Battery	19/04/2023	19/04/2023
							F1905	Device Revision or Replacement	A0705	Battery Problem	19/04/2023	19/04/2023
									A070504	Premature Discharge of Battery	19/04/2023	19/04/2023
					CT2406	Implantable	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem	19/04/2023	19/04/2023
									A070504	Premature Discharge of Battery	19/04/2023	19/04/2023
							F08	Hospitalization or Prolonged Hospitalization	A0705	Battery Problem	19/04/2023	19/04/2023
									A070504	Premature Discharge of Battery	19/04/2023	19/04/2023
							F1905	Device Revision or Replacement	A0705	Battery Problem	19/04/2023	19/04/2023
									A070504	Premature Discharge of Battery	19/04/2023	19/04/2023
					CT2407	Active implantable	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem	19/04/2023	19/04/2023
									A070504	Premature Discharge of Battery	19/04/2023	19/04/2023
							F08	Hospitalization or Prolonged Hospitalization	A0705	Battery Problem	19/04/2023	19/04/2023
									A070504	Premature Discharge of Battery	19/04/2023	19/04/2023
							F1905	Device Revision or Replacement	A0705	Battery Problem	19/04/2023	19/04/2023
									A070504	Premature Discharge of Battery	19/04/2023	19/04/2023
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem	19/04/2023	19/04/2023
									A070504	Premature Discharge of Battery	19/04/2023	19/04/2023
							F08	Hospitalization or Prolonged Hospitalization	A0705	Battery Problem	19/04/2023	19/04/2023
									A070504	Premature Discharge of Battery	19/04/2023	19/04/2023
							F1905	Device Revision or Replacement	A0705	Battery Problem	19/04/2023	19/04/2023
									A070504	Premature Discharge of Battery	19/04/2023	19/04/2023
					CT2727	Active implant devices/ancillaries	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem	19/04/2023	19/04/2023
									A070504	Premature Discharge of Battery	19/04/2023	19/04/2023
							F08	Hospitalization or Prolonged Hospitalization	A0705	Battery Problem	19/04/2023	19/04/2023
									A070504	Premature Discharge of Battery	19/04/2023	19/04/2023
							F1905	Device Revision or Replacement	A0705	Battery Problem	19/04/2023	19/04/2023
									A070504	Premature Discharge of Battery	19/04/2023	19/04/2023
					CT2750	Implantable cardiac electrical stimulators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem	19/04/2023	19/04/2023
									A070504	Premature Discharge of Battery	19/04/2023	19/04/2023
							F08	Hospitalization or Prolonged Hospitalization	A0705	Battery Problem	19/04/2023	19/04/2023
									A070504	Premature Discharge of Battery	19/04/2023	19/04/2023
							F1905	Device Revision or Replacement	A0705	Battery Problem	19/04/2023	19/04/2023
									A070504	Premature Discharge of Battery	19/04/2023	19/04/2023
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem	19/04/2023	19/04/2023
									A070504	Premature Discharge of Battery	19/04/2023	19/04/2023
							F08	Hospitalization or Prolonged Hospitalization	A0705	Battery Problem	19/04/2023	19/04/2023
									A070504	Premature Discharge of Battery	19/04/2023	19/04/2023
							F1905	Device Revision or Replacement	A0705	Battery Problem	19/04/2023	19/04/2023
									A070504	Premature Discharge of Battery	19/04/2023	19/04/2023
					CT2985	Active	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem	19/04/2023	19/04/2023
									A070504	Premature Discharge of Battery	19/04/2023	19/04/2023
							F08	Hospitalization or Prolonged Hospitalization	A0705	Battery Problem	19/04/2023	19/04/2023
									A070504	Premature Discharge of Battery	19/04/2023	19/04/2023
							F1905	Device Revision or Replacement	A0705	Battery Problem	19/04/2023	19/04/2023
									A070504	Premature Discharge of Battery	19/04/2023	19/04/2023
					CT3001	Implantable metal	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem	19/04/2023	19/04/2023
									A070504	Premature Discharge of Battery	19/04/2023	19/04/2023
							F08	Hospitalization or Prolonged Hospitalization	A0705	Battery Problem	19/04/2023	19/04/2023
									A070504	Premature Discharge of Battery	19/04/2023	19/04/2023
							F1905	Device Revision or Replacement	A0705	Battery Problem	19/04/2023	19/04/2023

								A070504	Premature Discharge of Battery	19/04/2023	19/04/2023
				CT3002	Implantable synthetic polymer	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem	19/04/2023	19/04/2023
								A070504	Premature Discharge of Battery	19/04/2023	19/04/2023
						F08	Hospitalization or Prolonged Hospitalization	A0705	Battery Problem	19/04/2023	19/04/2023
								A070504	Premature Discharge of Battery	19/04/2023	19/04/2023
						F1905	Device Revision or Replacement	A0705	Battery Problem	19/04/2023	19/04/2023
								A070504	Premature Discharge of Battery	19/04/2023	19/04/2023
				CT3006	Implantable inorganic material	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem	19/04/2023	19/04/2023
								A070504	Premature Discharge of Battery	19/04/2023	19/04/2023
						F08	Hospitalization or Prolonged Hospitalization	A0705	Battery Problem	19/04/2023	19/04/2023
								A070504	Premature Discharge of Battery	19/04/2023	19/04/2023
						F1905	Device Revision or Replacement	A0705	Battery Problem	19/04/2023	19/04/2023
								A070504	Premature Discharge of Battery	19/04/2023	19/04/2023
				CT314	Electrically powered	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem	19/04/2023	19/04/2023
								A070504	Premature Discharge of Battery	19/04/2023	19/04/2023
						F08	Hospitalization or Prolonged Hospitalization	A0705	Battery Problem	19/04/2023	19/04/2023
								A070504	Premature Discharge of Battery	19/04/2023	19/04/2023
						F1905	Device Revision or Replacement	A0705	Battery Problem	19/04/2023	19/04/2023
								A070504	Premature Discharge of Battery	19/04/2023	19/04/2023
				CT321	Long-term surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem	19/04/2023	19/04/2023
								A070504	Premature Discharge of Battery	19/04/2023	19/04/2023
						F08	Hospitalization or Prolonged Hospitalization	A0705	Battery Problem	19/04/2023	19/04/2023
								A070504	Premature Discharge of Battery	19/04/2023	19/04/2023
						F1905	Device Revision or Replacement	A0705	Battery Problem	19/04/2023	19/04/2023
								A070504	Premature Discharge of Battery	19/04/2023	19/04/2023
				CT449	Generators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem	19/04/2023	19/04/2023
								A070504	Premature Discharge of Battery	19/04/2023	19/04/2023
						F08	Hospitalization or Prolonged Hospitalization	A0705	Battery Problem	19/04/2023	19/04/2023
								A070504	Premature Discharge of Battery	19/04/2023	19/04/2023
						F1905	Device Revision or Replacement	A0705	Battery Problem	19/04/2023	19/04/2023
								A070504	Premature Discharge of Battery	19/04/2023	19/04/2023
				CT981	Single use	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem	19/04/2023	19/04/2023
								A070504	Premature Discharge of Battery	19/04/2023	19/04/2023
						F08	Hospitalization or Prolonged Hospitalization	A0705	Battery Problem	19/04/2023	19/04/2023
								A070504	Premature Discharge of Battery	19/04/2023	19/04/2023
						F1905	Device Revision or Replacement	A0705	Battery Problem	19/04/2023	19/04/2023
								A070504	Premature Discharge of Battery	19/04/2023	19/04/2023
				CT983	Surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem	19/04/2023	19/04/2023
								A070504	Premature Discharge of Battery	19/04/2023	19/04/2023
						F08	Hospitalization or Prolonged Hospitalization	A0705	Battery Problem	19/04/2023	19/04/2023
								A070504	Premature Discharge of Battery	19/04/2023	19/04/2023
						F1905	Device Revision or Replacement	A0705	Battery Problem	19/04/2023	19/04/2023
								A070504	Premature Discharge of Battery	19/04/2023	19/04/2023
13582	2023	All other reportable incidents	35223	Endocardial/interventricular septal pacing lead	CT1684	Cardiac pacemakers and associated devices	E0112	Dizziness	A070909	Oversensing		
									A0712	Pacing Problem		
							E011903	Syncope/Fainting	A070909	Oversensing		
									A0712	Pacing Problem		
							E0601	Arrhythmia	A070909	Oversensing		
									A0712	Pacing Problem		
							F1203	Life Threatening Illness or Injury	A070909	Oversensing		
									A0712	Pacing Problem		
13583	2023	All other reportable incidents	35223	Endocardial/interventricular septal pacing lead	CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070909	Oversensing	05/03/2020	05/03/2020
							F1901	Additional Surgery	A070909	Oversensing	05/03/2020	05/03/2020
13584	2023	All other reportable incidents	47265	Dual-chamber implantable pacemaker, rate-responsive	CT141	Electrodes/Leads and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem		
							F1901	Additional Surgery	A24	No Apparent Device Problem		
					CT145	Cardiovascular devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem		
							F1901	Additional Surgery	A24	No Apparent Device Problem		
					CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem		
							F1901	Additional Surgery	A24	No Apparent Device Problem		
					CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem		
							F1901	Additional Surgery	A24	No Apparent Device Problem		
					CT233	Surgical	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem		
							F1901	Additional Surgery	A24	No Apparent Device Problem		
					CT2406	Implantable	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem		
							F1901	Additional Surgery	A24	No Apparent Device Problem		
					CT2407	Active implantable	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem		
							F1901	Additional Surgery	A24	No Apparent Device Problem		
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem		
							F1901	Additional Surgery	A24	No Apparent Device Problem		
					CT2727	Active implant devices/ancillaries	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem		
							F1901	Additional Surgery	A24	No Apparent Device Problem		
					CT2750	Implantable cardiac electrical stimulators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem		
							F1901	Additional Surgery	A24	No Apparent Device Problem		
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem		
							F1901	Additional Surgery	A24	No Apparent Device Problem		
					CT2985	Active	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem		
							F1901	Additional Surgery	A24	No Apparent Device Problem		
					CT3001	Implantable metal	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem		
							F1901	Additional Surgery	A24	No Apparent Device Problem		
					CT3002	Implantable synthetic polymer	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem		
							F1901	Additional Surgery	A24	No Apparent Device Problem		
					CT3006	Implantable inorganic material	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem		
							F1901	Additional Surgery	A24	No Apparent Device Problem		
					CT314	Electrically powered	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem		
							F1901	Additional Surgery	A24	No Apparent Device Problem		
					CT321	Long-term surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem		
							F1901	Additional Surgery	A24	No Apparent Device Problem		
					CT449	Generators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem		

						F1901	Additional Surgery	A24	No Apparent Device Problem		
					CT981	Single use	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem		
							F1901	Additional Surgery	A24	No Apparent Device Problem		
					CT983	Surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem		
							F1901	Additional Surgery	A24	No Apparent Device Problem		
13585	2023	All other reportable incidents	47263	Cardiac resynchronization therapy implantable pacemaker	CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A0908	Incorrect, Inadequate or Imprecise Result or Readings	31/07/2023	31/07/2023
							F08	Hospitalization or Prolonged Hospitalization	A0908	Incorrect, Inadequate or Imprecise Result or Readings	31/07/2023	31/07/2023
							F1905	Device Revision or Replacement	A0908	Incorrect, Inadequate or Imprecise Result or Readings	31/07/2023	31/07/2023
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A0908	Incorrect, Inadequate or Imprecise Result or Readings	31/07/2023	31/07/2023
							F08	Hospitalization or Prolonged Hospitalization	A0908	Incorrect, Inadequate or Imprecise Result or Readings	31/07/2023	31/07/2023
							F1905	Device Revision or Replacement	A0908	Incorrect, Inadequate or Imprecise Result or Readings	31/07/2023	31/07/2023
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A0908	Incorrect, Inadequate or Imprecise Result or Readings	31/07/2023	31/07/2023
							F08	Hospitalization or Prolonged Hospitalization	A0908	Incorrect, Inadequate or Imprecise Result or Readings	31/07/2023	31/07/2023
							F1905	Device Revision or Replacement	A0908	Incorrect, Inadequate or Imprecise Result or Readings	31/07/2023	31/07/2023
13586	2023	All other reportable incidents	35223	Endocardial/interventricular septal pacing lead	CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070901	Decreased Sensitivity		
									A070909	Oversensing		
							F08	Hospitalization or Prolonged Hospitalization	A070901	Decreased Sensitivity		
									A070909	Oversensing		
							F12	Serious Injury/ Illness/ Impairment	A070901	Decreased Sensitivity		
									A070909	Oversensing		
							F23	Unexpected Medical Intervention	A070901	Decreased Sensitivity		
									A070909	Oversensing		
13587	2023	All other reportable incidents	47265	Dual-chamber implantable pacemaker, rate-responsive	CT141	Electrodes/Leads and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
									A090811	Unable to Obtain Readings		
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery		
									A090811	Unable to Obtain Readings		
					CT145	Cardiovascular devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
									A090811	Unable to Obtain Readings		
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery		
									A090811	Unable to Obtain Readings		
					CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
									A090811	Unable to Obtain Readings		
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery		
									A090811	Unable to Obtain Readings		
					CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
									A090811	Unable to Obtain Readings		
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery		
									A090811	Unable to Obtain Readings		
					CT233	Surgical	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
									A090811	Unable to Obtain Readings		
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery		
									A090811	Unable to Obtain Readings		
					CT2406	Implantable	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
									A090811	Unable to Obtain Readings		
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery		
									A090811	Unable to Obtain Readings		
					CT2407	Active implantable	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
									A090811	Unable to Obtain Readings		
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery		
									A090811	Unable to Obtain Readings		
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
									A090811	Unable to Obtain Readings		
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery		
									A090811	Unable to Obtain Readings		
					CT2727	Active implant devices/ancillaries	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
									A090811	Unable to Obtain Readings		
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery		
									A090811	Unable to Obtain Readings		
					CT2750	Implantable cardiac electrical stimulators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
									A090811	Unable to Obtain Readings		
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery		
									A090811	Unable to Obtain Readings		
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
									A090811	Unable to Obtain Readings		
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery		
									A090811	Unable to Obtain Readings		
					CT2985	Active	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
									A090811	Unable to Obtain Readings		
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery		
									A090811	Unable to Obtain Readings		
					CT3001	Implantable metal	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
									A090811	Unable to Obtain Readings		
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery		
									A090811	Unable to Obtain Readings		
					CT3002	Implantable synthetic polymer	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
									A090811	Unable to Obtain Readings		
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery		
									A090811	Unable to Obtain Readings		
					CT3006	Implantable inorganic material	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
									A090811	Unable to Obtain Readings		
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery		

									A090811	Unable to Obtain Readings		
					CT314	Electrically powered	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
									A090811	Unable to Obtain Readings		
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery		
									A090811	Unable to Obtain Readings		
					CT321	Long-term surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
									A090811	Unable to Obtain Readings		
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery		
									A090811	Unable to Obtain Readings		
					CT449	Generators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
									A090811	Unable to Obtain Readings		
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery		
									A090811	Unable to Obtain Readings		
					CT981	Single use	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
									A090811	Unable to Obtain Readings		
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery		
									A090811	Unable to Obtain Readings		
					CT983	Surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
									A090811	Unable to Obtain Readings		
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery		
									A090811	Unable to Obtain Readings		
13588	2023	All other reportable incidents	47265	Dual-chamber implantable pacemaker, rate-responsive	CT141	Electrodes/Leads and associated devices	E011902	Presyncope	A0712	Pacing Problem		
							F12	Serious Injury/ Illness/ Impairment	A0712	Pacing Problem		
							F1905	Device Revision or Replacement	A0712	Pacing Problem		
							F23	Unexpected Medical Intervention	A0712	Pacing Problem		
					CT145	Cardiovascular devices	E011902	Presyncope	A0712	Pacing Problem		
							F12	Serious Injury/ Illness/ Impairment	A0712	Pacing Problem		
							F1905	Device Revision or Replacement	A0712	Pacing Problem		
							F23	Unexpected Medical Intervention	A0712	Pacing Problem		
					CT1684	Cardiac pacemakers and associated devices	E011902	Presyncope	A0712	Pacing Problem		
							F12	Serious Injury/ Illness/ Impairment	A0712	Pacing Problem		
							F1905	Device Revision or Replacement	A0712	Pacing Problem		
							F23	Unexpected Medical Intervention	A0712	Pacing Problem		
					CT1685	Cardiac pacemakers	E011902	Presyncope	A0712	Pacing Problem		
							F12	Serious Injury/ Illness/ Impairment	A0712	Pacing Problem		
							F1905	Device Revision or Replacement	A0712	Pacing Problem		
							F23	Unexpected Medical Intervention	A0712	Pacing Problem		
					CT233	Surgical	E011902	Presyncope	A0712	Pacing Problem		
							F12	Serious Injury/ Illness/ Impairment	A0712	Pacing Problem		
							F1905	Device Revision or Replacement	A0712	Pacing Problem		
							F23	Unexpected Medical Intervention	A0712	Pacing Problem		
					CT2406	Implantable	E011902	Presyncope	A0712	Pacing Problem		
							F12	Serious Injury/ Illness/ Impairment	A0712	Pacing Problem		
							F1905	Device Revision or Replacement	A0712	Pacing Problem		
							F23	Unexpected Medical Intervention	A0712	Pacing Problem		
					CT2407	Active implantable	E011902	Presyncope	A0712	Pacing Problem		
							F12	Serious Injury/ Illness/ Impairment	A0712	Pacing Problem		
							F1905	Device Revision or Replacement	A0712	Pacing Problem		
							F23	Unexpected Medical Intervention	A0712	Pacing Problem		
					CT257	Implantable cardiac pacemakers	E011902	Presyncope	A0712	Pacing Problem		
							F12	Serious Injury/ Illness/ Impairment	A0712	Pacing Problem		
							F1905	Device Revision or Replacement	A0712	Pacing Problem		
							F23	Unexpected Medical Intervention	A0712	Pacing Problem		
					CT2727	Active implant devices/ancillaries	E011902	Presyncope	A0712	Pacing Problem		
							F12	Serious Injury/ Illness/ Impairment	A0712	Pacing Problem		
							F1905	Device Revision or Replacement	A0712	Pacing Problem		
							F23	Unexpected Medical Intervention	A0712	Pacing Problem		
					CT2750	Implantable cardiac electrical stimulators and associated devices	E011902	Presyncope	A0712	Pacing Problem		
							F12	Serious Injury/ Illness/ Impairment	A0712	Pacing Problem		
							F1905	Device Revision or Replacement	A0712	Pacing Problem		
							F23	Unexpected Medical Intervention	A0712	Pacing Problem		
					CT2866	Implantable cardiac electrical stimulators	E011902	Presyncope	A0712	Pacing Problem		
							F12	Serious Injury/ Illness/ Impairment	A0712	Pacing Problem		
							F1905	Device Revision or Replacement	A0712	Pacing Problem		
							F23	Unexpected Medical Intervention	A0712	Pacing Problem		
					CT2985	Active	E011902	Presyncope	A0712	Pacing Problem		
							F12	Serious Injury/ Illness/ Impairment	A0712	Pacing Problem		
							F1905	Device Revision or Replacement	A0712	Pacing Problem		
							F23	Unexpected Medical Intervention	A0712	Pacing Problem		
					CT3001	Implantable metal	E011902	Presyncope	A0712	Pacing Problem		
							F12	Serious Injury/ Illness/ Impairment	A0712	Pacing Problem		
							F1905	Device Revision or Replacement	A0712	Pacing Problem		
							F23	Unexpected Medical Intervention	A0712	Pacing Problem		
					CT3002	Implantable synthetic polymer	E011902	Presyncope	A0712	Pacing Problem		
							F12	Serious Injury/ Illness/ Impairment	A0712	Pacing Problem		
							F1905	Device Revision or Replacement	A0712	Pacing Problem		
							F23	Unexpected Medical Intervention	A0712	Pacing Problem		
					CT3006	Implantable inorganic material	E011902	Presyncope	A0712	Pacing Problem		
							F12	Serious Injury/ Illness/ Impairment	A0712	Pacing Problem		
							F1905	Device Revision or Replacement	A0712	Pacing Problem		
							F23	Unexpected Medical Intervention	A0712	Pacing Problem		
					CT314	Electrically powered	E011902	Presyncope	A0712	Pacing Problem		
							F12	Serious Injury/ Illness/ Impairment	A0712	Pacing Problem		
							F1905	Device Revision or Replacement	A0712	Pacing Problem		
							F23	Unexpected Medical Intervention	A0712	Pacing Problem		
					CT321	Long-term surgical invasive	E011902	Presyncope	A0712	Pacing Problem		
							F12	Serious Injury/ Illness/ Impairment	A0712	Pacing Problem		
							F1905	Device Revision or Replacement	A0712	Pacing Problem		

							F23	Unexpected Medical Intervention	A0712	Pacing Problem		
					CT449	Generators and associated devices	E011902	Presyncope	A0712	Pacing Problem		
							F12	Serious Injury/ Illness/ Impairment	A0712	Pacing Problem		
							F1905	Device Revision or Replacement	A0712	Pacing Problem		
							F23	Unexpected Medical Intervention	A0712	Pacing Problem		
					CT981	Single use	E011902	Presyncope	A0712	Pacing Problem		
							F12	Serious Injury/ Illness/ Impairment	A0712	Pacing Problem		
							F1905	Device Revision or Replacement	A0712	Pacing Problem		
							F23	Unexpected Medical Intervention	A0712	Pacing Problem		
					CT983	Surgical invasive	E011902	Presyncope	A0712	Pacing Problem		
							F12	Serious Injury/ Illness/ Impairment	A0712	Pacing Problem		
							F1905	Device Revision or Replacement	A0712	Pacing Problem		
							F23	Unexpected Medical Intervention	A0712	Pacing Problem		
13589	2023	All other reportable incidents	47265	Dual-chamber implantable pacemaker, rate-responsive	CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
13590	2023		47265	Dual-chamber implantable pacemaker, rate-responsive	CT141	Electrodes/Leads and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	27/06/2023	27/06/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	27/06/2023	27/06/2023
					CT145	Cardiovascular devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	27/06/2023	27/06/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	27/06/2023	27/06/2023
					CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	27/06/2023	27/06/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	27/06/2023	27/06/2023
					CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	27/06/2023	27/06/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	27/06/2023	27/06/2023
					CT233	Surgical	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	27/06/2023	27/06/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	27/06/2023	27/06/2023
					CT2406	Implantable	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	27/06/2023	27/06/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	27/06/2023	27/06/2023
					CT2407	Active implantable	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	27/06/2023	27/06/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	27/06/2023	27/06/2023
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	27/06/2023	27/06/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	27/06/2023	27/06/2023
					CT2727	Active implant devices/ancillaries	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	27/06/2023	27/06/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	27/06/2023	27/06/2023
					CT2750	Implantable cardiac electrical stimulators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	27/06/2023	27/06/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	27/06/2023	27/06/2023
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	27/06/2023	27/06/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	27/06/2023	27/06/2023
					CT2985	Active	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	27/06/2023	27/06/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	27/06/2023	27/06/2023
					CT3001	Implantable metal	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	27/06/2023	27/06/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	27/06/2023	27/06/2023
					CT3002	Implantable synthetic polymer	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	27/06/2023	27/06/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	27/06/2023	27/06/2023
					CT3006	Implantable inorganic material	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	27/06/2023	27/06/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	27/06/2023	27/06/2023
					CT314	Electrically powered	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	27/06/2023	27/06/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	27/06/2023	27/06/2023
					CT321	Long-term surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	27/06/2023	27/06/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	27/06/2023	27/06/2023
					CT449	Generators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	27/06/2023	27/06/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	27/06/2023	27/06/2023
					CT981	Single use	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	27/06/2023	27/06/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	27/06/2023	27/06/2023
					CT983	Surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	27/06/2023	27/06/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	27/06/2023	27/06/2023
13591	2023		47265	Dual-chamber implantable pacemaker, rate-responsive	CT141	Electrodes/Leads and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	03/07/2023	03/07/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	03/07/2023	03/07/2023
					CT145	Cardiovascular devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	03/07/2023	03/07/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	03/07/2023	03/07/2023
					CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	03/07/2023	03/07/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	03/07/2023	03/07/2023
					CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	03/07/2023	03/07/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	03/07/2023	03/07/2023
					CT233	Surgical	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	03/07/2023	03/07/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	03/07/2023	03/07/2023
					CT2406	Implantable	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	03/07/2023	03/07/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	03/07/2023	03/07/2023
					CT2407	Active implantable	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	03/07/2023	03/07/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	03/07/2023	03/07/2023
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	03/07/2023	03/07/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	03/07/2023	03/07/2023
					CT2727	Active implant devices/ancillaries	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	03/07/2023	03/07/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	03/07/2023	03/07/2023
					CT2750	Implantable cardiac electrical stimulators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	03/07/2023	03/07/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	03/07/2023	03/07/2023
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	03/07/2023	03/07/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	03/07/2023	03/07/2023
					CT2985	Active	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	03/07/2023	03/07/2023

							F1901	Additional Surgery	A24	No Apparent Device Problem	03/07/2023	03/07/2023
					CT3001	Implantable metal	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	03/07/2023	03/07/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	03/07/2023	03/07/2023
					CT3002	Implantable synthetic polymer	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	03/07/2023	03/07/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	03/07/2023	03/07/2023
					CT3006	Implantable inorganic material	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	03/07/2023	03/07/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	03/07/2023	03/07/2023
					CT314	Electrically powered	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	03/07/2023	03/07/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	03/07/2023	03/07/2023
					CT321	Long-term surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	03/07/2023	03/07/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	03/07/2023	03/07/2023
					CT449	Generators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	03/07/2023	03/07/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	03/07/2023	03/07/2023
					CT981	Single use	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	03/07/2023	03/07/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	03/07/2023	03/07/2023
					CT983	Surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	03/07/2023	03/07/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	03/07/2023	03/07/2023
13592	2023		47265	Dual-chamber implantable pacemaker, rate-responsive	CT141	Electrodes/Leads and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	14/07/2023	14/07/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	14/07/2023	14/07/2023
					CT145	Cardiovascular devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	14/07/2023	14/07/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	14/07/2023	14/07/2023
					CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	14/07/2023	14/07/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	14/07/2023	14/07/2023
					CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	14/07/2023	14/07/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	14/07/2023	14/07/2023
					CT233	Surgical	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	14/07/2023	14/07/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	14/07/2023	14/07/2023
					CT2406	Implantable	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	14/07/2023	14/07/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	14/07/2023	14/07/2023
					CT2407	Active implantable	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	14/07/2023	14/07/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	14/07/2023	14/07/2023
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	14/07/2023	14/07/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	14/07/2023	14/07/2023
					CT2727	Active implant devices/ancillaries	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	14/07/2023	14/07/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	14/07/2023	14/07/2023
					CT2750	Implantable cardiac electrical stimulators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	14/07/2023	14/07/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	14/07/2023	14/07/2023
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	14/07/2023	14/07/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	14/07/2023	14/07/2023
					CT2985	Active	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	14/07/2023	14/07/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	14/07/2023	14/07/2023
					CT3001	Implantable metal	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	14/07/2023	14/07/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	14/07/2023	14/07/2023
					CT3002	Implantable synthetic polymer	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	14/07/2023	14/07/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	14/07/2023	14/07/2023
					CT3006	Implantable inorganic material	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	14/07/2023	14/07/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	14/07/2023	14/07/2023
					CT314	Electrically powered	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	14/07/2023	14/07/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	14/07/2023	14/07/2023
					CT321	Long-term surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	14/07/2023	14/07/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	14/07/2023	14/07/2023
					CT449	Generators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	14/07/2023	14/07/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	14/07/2023	14/07/2023
					CT981	Single use	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	14/07/2023	14/07/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	14/07/2023	14/07/2023
					CT983	Surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	14/07/2023	14/07/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	14/07/2023	14/07/2023
13593	2023	All other reportable incidents	47270	Cardiac resynchronization therapy implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E0601	Arrhythmia	A0709	Device Sensing Problem		
									A070909	Oversensing		
									A070910	Undersensing		
									A071302	Inappropriate/Inadequate Shock/Stimulation		
							F12	Serious Injury/ Illness/ Impairment	A0709	Device Sensing Problem		
									A070909	Oversensing		
									A070910	Undersensing		
									A071302	Inappropriate/Inadequate Shock/Stimulation		
13594	2023	All other reportable incidents	37265	Dual-chamber implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A04	Material Integrity Problem	06/03/2023	06/03/2023
									A070909	Oversensing	06/03/2023	06/03/2023
									A090801	Signal Artifact/Noise	06/03/2023	06/03/2023
							F08	Hospitalization or Prolonged Hospitalization	A04	Material Integrity Problem	06/03/2023	06/03/2023
									A070909	Oversensing	06/03/2023	06/03/2023
									A090801	Signal Artifact/Noise	06/03/2023	06/03/2023
							F1905	Device Revision or Replacement	A04	Material Integrity Problem	06/03/2023	06/03/2023
									A070909	Oversensing	06/03/2023	06/03/2023
									A090801	Signal Artifact/Noise	06/03/2023	06/03/2023
13595	2023	All other reportable incidents	47265	Dual-chamber implantable pacemaker, rate-responsive	CT141	Electrodes/Leads and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT145	Cardiovascular devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT233	Surgical	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT2406	Implantable	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT2407	Active implantable	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		

					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT2727	Active implant devices/ancillaries	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT2750	Implantable cardiac electrical stimulators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT2985	Active	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT3001	Implantable metal	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT3002	Implantable synthetic polymer	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT3006	Implantable inorganic material	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT314	Electrically powered	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT321	Long-term surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT449	Generators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT981	Single use	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT983	Surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
13596	2023	All other reportable incidents	17882	Physiologic-monitoring defibrillation system	CT1684	Cardiac pacemakers and associated devices	E24	Other Terms/Codes Related to Clinical Signs and Symptoms or Conditions	A0705	Battery Problem		
							F27	Problem Identified During Non-Clinical Procedure	A0705	Battery Problem		
13597	2023	All other reportable incidents	47263	Cardiac resynchronization therapy implantable pacemaker	CT1684	Cardiac pacemakers and associated devices	E2103	Device Overstimulation of Tissue	A0908	Incorrect, Inadequate or Imprecise Result or Readings	16/01/2023	16/01/2023
							F08	Hospitalization or Prolonged Hospitalization	A0908	Incorrect, Inadequate or Imprecise Result or Readings	16/01/2023	16/01/2023
							F1905	Device Revision or Replacement	A0908	Incorrect, Inadequate or Imprecise Result or Readings	16/01/2023	16/01/2023
					CT1685	Cardiac pacemakers	E2103	Device Overstimulation of Tissue	A0908	Incorrect, Inadequate or Imprecise Result or Readings	16/01/2023	16/01/2023
							F08	Hospitalization or Prolonged Hospitalization	A0908	Incorrect, Inadequate or Imprecise Result or Readings	16/01/2023	16/01/2023
							F1905	Device Revision or Replacement	A0908	Incorrect, Inadequate or Imprecise Result or Readings	16/01/2023	16/01/2023
					CT257	Implantable cardiac pacemakers	E2103	Device Overstimulation of Tissue	A0908	Incorrect, Inadequate or Imprecise Result or Readings	16/01/2023	16/01/2023
							F08	Hospitalization or Prolonged Hospitalization	A0908	Incorrect, Inadequate or Imprecise Result or Readings	16/01/2023	16/01/2023
							F1905	Device Revision or Replacement	A0908	Incorrect, Inadequate or Imprecise Result or Readings	16/01/2023	16/01/2023
					CT2866	Implantable cardiac electrical stimulators	E2103	Device Overstimulation of Tissue	A0908	Incorrect, Inadequate or Imprecise Result or Readings	16/01/2023	16/01/2023
							F08	Hospitalization or Prolonged Hospitalization	A0908	Incorrect, Inadequate or Imprecise Result or Readings	16/01/2023	16/01/2023
							F1905	Device Revision or Replacement	A0908	Incorrect, Inadequate or Imprecise Result or Readings	16/01/2023	16/01/2023
13598	2023	All other reportable incidents	47267	Single-chamber implantable pacemaker, rate-responsive	CT141	Electrodes/Leads and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	16/03/2023	16/03/2023
							F1901	Additional Surgery	A070504	Premature Discharge of Battery	16/03/2023	16/03/2023
					CT145	Cardiovascular devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	16/03/2023	16/03/2023
							F1901	Additional Surgery	A070504	Premature Discharge of Battery	16/03/2023	16/03/2023
					CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	16/03/2023	16/03/2023
							F1901	Additional Surgery	A070504	Premature Discharge of Battery	16/03/2023	16/03/2023
					CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	16/03/2023	16/03/2023
							F1901	Additional Surgery	A070504	Premature Discharge of Battery	16/03/2023	16/03/2023
					CT233	Surgical	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	16/03/2023	16/03/2023
							F1901	Additional Surgery	A070504	Premature Discharge of Battery	16/03/2023	16/03/2023
					CT2406	Implantable	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	16/03/2023	16/03/2023
							F1901	Additional Surgery	A070504	Premature Discharge of Battery	16/03/2023	16/03/2023
					CT2407	Active implantable	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	16/03/2023	16/03/2023
							F1901	Additional Surgery	A070504	Premature Discharge of Battery	16/03/2023	16/03/2023
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	16/03/2023	16/03/2023
							F1901	Additional Surgery	A070504	Premature Discharge of Battery	16/03/2023	16/03/2023
					CT2727	Active implant devices/ancillaries	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	16/03/2023	16/03/2023
							F1901	Additional Surgery	A070504	Premature Discharge of Battery	16/03/2023	16/03/2023
					CT2750	Implantable cardiac electrical stimulators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	16/03/2023	16/03/2023
							F1901	Additional Surgery	A070504	Premature Discharge of Battery	16/03/2023	16/03/2023
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	16/03/2023	16/03/2023
							F1901	Additional Surgery	A070504	Premature Discharge of Battery	16/03/2023	16/03/2023
					CT2985	Active	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	16/03/2023	16/03/2023
							F1901	Additional Surgery	A070504	Premature Discharge of Battery	16/03/2023	16/03/2023
					CT3001	Implantable metal	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	16/03/2023	16/03/2023
							F1901	Additional Surgery	A070504	Premature Discharge of Battery	16/03/2023	16/03/2023
					CT3002	Implantable synthetic polymer	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	16/03/2023	16/03/2023
							F1901	Additional Surgery	A070504	Premature Discharge of Battery	16/03/2023	16/03/2023
					CT3006	Implantable inorganic material	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	16/03/2023	16/03/2023
							F1901	Additional Surgery	A070504	Premature Discharge of Battery	16/03/2023	16/03/2023
					CT314	Electrically powered	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	16/03/2023	16/03/2023
							F1901	Additional Surgery	A070504	Premature Discharge of Battery	16/03/2023	16/03/2023
					CT321	Long-term surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	16/03/2023	16/03/2023
							F1901	Additional Surgery	A070504	Premature Discharge of Battery	16/03/2023	16/03/2023
					CT449	Generators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	16/03/2023	16/03/2023
							F1901	Additional Surgery	A070504	Premature Discharge of Battery	16/03/2023	16/03/2023
					CT981	Single use	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	16/03/2023	16/03/2023
							F1901	Additional Surgery	A070504	Premature Discharge of Battery	16/03/2023	16/03/2023

					CT983	Surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	16/03/2023	16/03/2023
							F1901	Additional Surgery	A070504	Premature Discharge of Battery	16/03/2023	16/03/2023
13599	2023	All other reportable incidents	35852	Single-chamber implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A071102	Failure to Interrogate		
									A160503	Inappropriate or Unexpected Reset		
							F1901	Additional Surgery	A071102	Failure to Interrogate		
									A160503	Inappropriate or Unexpected Reset		
13600	2023	All other reportable incidents	47265	Dual-chamber implantable pacemaker, rate-responsive	CT141	Electrodes/Leads and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT145	Cardiovascular devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT233	Surgical	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT2406	Implantable	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT2407	Active implantable	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT2727	Active implant devices/ancillaries	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT2750	Implantable cardiac electrical stimulators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT2985	Active	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT3001	Implantable metal	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT3002	Implantable synthetic polymer	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT3006	Implantable inorganic material	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT314	Electrically powered	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT321	Long-term surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT449	Generators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT981	Single use	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT983	Surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
13601	2023		47265	Dual-chamber implantable pacemaker, rate-responsive	CT141	Electrodes/Leads and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	17/06/2023	17/06/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	17/06/2023	17/06/2023
					CT145	Cardiovascular devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	17/06/2023	17/06/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	17/06/2023	17/06/2023
					CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	17/06/2023	17/06/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	17/06/2023	17/06/2023
					CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	17/06/2023	17/06/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	17/06/2023	17/06/2023
					CT233	Surgical	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	17/06/2023	17/06/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	17/06/2023	17/06/2023
					CT2406	Implantable	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	17/06/2023	17/06/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	17/06/2023	17/06/2023
					CT2407	Active implantable	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	17/06/2023	17/06/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	17/06/2023	17/06/2023
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	17/06/2023	17/06/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	17/06/2023	17/06/2023
					CT2727	Active implant devices/ancillaries	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	17/06/2023	17/06/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	17/06/2023	17/06/2023
					CT2750	Implantable cardiac electrical stimulators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	17/06/2023	17/06/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	17/06/2023	17/06/2023
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	17/06/2023	17/06/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	17/06/2023	17/06/2023
					CT2985	Active	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	17/06/2023	17/06/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	17/06/2023	17/06/2023
					CT3001	Implantable metal	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	17/06/2023	17/06/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	17/06/2023	17/06/2023
					CT3002	Implantable synthetic polymer	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	17/06/2023	17/06/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	17/06/2023	17/06/2023
					CT3006	Implantable inorganic material	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	17/06/2023	17/06/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	17/06/2023	17/06/2023
					CT314	Electrically powered	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	17/06/2023	17/06/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	17/06/2023	17/06/2023
					CT321	Long-term surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	17/06/2023	17/06/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	17/06/2023	17/06/2023
					CT449	Generators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	17/06/2023	17/06/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	17/06/2023	17/06/2023
					CT981	Single use	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	17/06/2023	17/06/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	17/06/2023	17/06/2023
					CT983	Surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	17/06/2023	17/06/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	17/06/2023	17/06/2023
13602	2023		47265	Dual-chamber implantable pacemaker, rate-responsive	CT141	Electrodes/Leads and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	17/06/2023	17/06/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	17/06/2023	17/06/2023

					CT145	Cardiovascular devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	17/06/2023	17/06/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	17/06/2023	17/06/2023
					CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	17/06/2023	17/06/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	17/06/2023	17/06/2023
					CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	17/06/2023	17/06/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	17/06/2023	17/06/2023
					CT233	Surgical	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	17/06/2023	17/06/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	17/06/2023	17/06/2023
					CT2406	Implantable	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	17/06/2023	17/06/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	17/06/2023	17/06/2023
					CT2407	Active implantable	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	17/06/2023	17/06/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	17/06/2023	17/06/2023
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	17/06/2023	17/06/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	17/06/2023	17/06/2023
					CT2727	Active implant devices/ancillaries	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	17/06/2023	17/06/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	17/06/2023	17/06/2023
					CT2750	Implantable cardiac electrical stimulators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	17/06/2023	17/06/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	17/06/2023	17/06/2023
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	17/06/2023	17/06/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	17/06/2023	17/06/2023
					CT2985	Active	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	17/06/2023	17/06/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	17/06/2023	17/06/2023
					CT3001	Implantable metal	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	17/06/2023	17/06/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	17/06/2023	17/06/2023
					CT3002	Implantable synthetic polymer	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	17/06/2023	17/06/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	17/06/2023	17/06/2023
					CT3006	Implantable inorganic material	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	17/06/2023	17/06/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	17/06/2023	17/06/2023
					CT314	Electrically powered	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	17/06/2023	17/06/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	17/06/2023	17/06/2023
					CT321	Long-term surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	17/06/2023	17/06/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	17/06/2023	17/06/2023
					CT449	Generators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	17/06/2023	17/06/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	17/06/2023	17/06/2023
					CT981	Single use	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	17/06/2023	17/06/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	17/06/2023	17/06/2023
					CT983	Surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	17/06/2023	17/06/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	17/06/2023	17/06/2023
13603	2023		47265	Dual-chamber implantable pacemaker, rate-responsive	CT141	Electrodes/Leads and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	17/06/2023	17/06/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	17/06/2023	17/06/2023
					CT145	Cardiovascular devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	17/06/2023	17/06/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	17/06/2023	17/06/2023
					CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	17/06/2023	17/06/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	17/06/2023	17/06/2023
					CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	17/06/2023	17/06/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	17/06/2023	17/06/2023
					CT233	Surgical	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	17/06/2023	17/06/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	17/06/2023	17/06/2023
					CT2406	Implantable	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	17/06/2023	17/06/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	17/06/2023	17/06/2023
					CT2407	Active implantable	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	17/06/2023	17/06/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	17/06/2023	17/06/2023
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	17/06/2023	17/06/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	17/06/2023	17/06/2023
					CT2727	Active implant devices/ancillaries	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	17/06/2023	17/06/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	17/06/2023	17/06/2023
					CT2750	Implantable cardiac electrical stimulators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	17/06/2023	17/06/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	17/06/2023	17/06/2023
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	17/06/2023	17/06/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	17/06/2023	17/06/2023
					CT2985	Active	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	17/06/2023	17/06/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	17/06/2023	17/06/2023
					CT3001	Implantable metal	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	17/06/2023	17/06/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	17/06/2023	17/06/2023
					CT3002	Implantable synthetic polymer	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	17/06/2023	17/06/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	17/06/2023	17/06/2023
					CT3006	Implantable inorganic material	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	17/06/2023	17/06/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	17/06/2023	17/06/2023
					CT314	Electrically powered	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	17/06/2023	17/06/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	17/06/2023	17/06/2023
					CT321	Long-term surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	17/06/2023	17/06/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	17/06/2023	17/06/2023
					CT449	Generators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	17/06/2023	17/06/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	17/06/2023	17/06/2023
					CT981	Single use	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	17/06/2023	17/06/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	17/06/2023	17/06/2023
					CT983	Surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	17/06/2023	17/06/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	17/06/2023	17/06/2023
13604	2023		47265	Dual-chamber implantable pacemaker, rate-responsive	CT141	Electrodes/Leads and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	17/06/2023	17/06/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	17/06/2023	17/06/2023
					CT145	Cardiovascular devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	17/06/2023	17/06/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	17/06/2023	17/06/2023
					CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	17/06/2023	17/06/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	17/06/2023	17/06/2023
					CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	17/06/2023	17/06/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	17/06/2023	17/06/2023
					CT233	Surgical	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	17/06/2023	17/06/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	17/06/2023	17/06/2023

					CT2406	Implantable	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	17/06/2023	17/06/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	17/06/2023	17/06/2023
					CT2407	Active implantable	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	17/06/2023	17/06/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	17/06/2023	17/06/2023
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	17/06/2023	17/06/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	17/06/2023	17/06/2023
					CT2727	Active implant devices/ancillaries	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	17/06/2023	17/06/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	17/06/2023	17/06/2023
					CT2750	Implantable cardiac electrical stimulators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	17/06/2023	17/06/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	17/06/2023	17/06/2023
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	17/06/2023	17/06/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	17/06/2023	17/06/2023
					CT2985	Active	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	17/06/2023	17/06/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	17/06/2023	17/06/2023
					CT3001	Implantable metal	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	17/06/2023	17/06/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	17/06/2023	17/06/2023
					CT3002	Implantable synthetic polymer	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	17/06/2023	17/06/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	17/06/2023	17/06/2023
					CT3006	Implantable inorganic material	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	17/06/2023	17/06/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	17/06/2023	17/06/2023
					CT314	Electrically powered	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	17/06/2023	17/06/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	17/06/2023	17/06/2023
					CT321	Long-term surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	17/06/2023	17/06/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	17/06/2023	17/06/2023
					CT449	Generators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	17/06/2023	17/06/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	17/06/2023	17/06/2023
					CT981	Single use	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	17/06/2023	17/06/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	17/06/2023	17/06/2023
					CT983	Surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	17/06/2023	17/06/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	17/06/2023	17/06/2023
13605	2023	All other reportable incidents	35223	Endocardial/interventricular septal pacing lead	CT1684	Cardiac pacemakers and associated devices	E0112	Dizziness	A070101	Failure to Capture		
									A070901	Decreased Sensitivity		
									A070909	Oversensing		
									A090801	Signal Artifact/Noise		
							E013403	Paresthesia	A070101	Failure to Capture		
									A070901	Decreased Sensitivity		
									A070909	Oversensing		
									A090801	Signal Artifact/Noise		
							E2338	Swelling/ Edema	A070101	Failure to Capture		
									A070901	Decreased Sensitivity		
									A070909	Oversensing		
									A090801	Signal Artifact/Noise		
							F12	Serious Injury/ Illness/ Impairment	A070101	Failure to Capture		
									A070901	Decreased Sensitivity		
									A070909	Oversensing		
									A090801	Signal Artifact/Noise		
							F23	Unexpected Medical Intervention	A070101	Failure to Capture		
									A070901	Decreased Sensitivity		
									A070909	Oversensing		
									A090801	Signal Artifact/Noise		
13606	2023	All other reportable incidents	47205	Cardiac pulse generator programmer	CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A090805	Incorrect Measurement		
							F23	Unexpected Medical Intervention	A090805	Incorrect Measurement		
13607	2023	All other reportable incidents	47205	Cardiac pulse generator programmer	CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A090805	Incorrect Measurement		
							F23	Unexpected Medical Intervention	A090805	Incorrect Measurement		
13608	2023	All other reportable incidents	47205	Cardiac pulse generator programmer	CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A090805	Incorrect Measurement		
							F23	Unexpected Medical Intervention	A090805	Incorrect Measurement		
13609	2023	All other reportable incidents	60789	Intracardiac pacemaker	CT130	Catheters and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A1502	Positioning Problem		
							F23	Unexpected Medical Intervention	A1502	Positioning Problem		
					CT145	Cardiovascular devices	E2403	No Clinical Signs, Symptoms or Conditions	A1502	Positioning Problem		
							F23	Unexpected Medical Intervention	A1502	Positioning Problem		
					CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A1502	Positioning Problem		
							F23	Unexpected Medical Intervention	A1502	Positioning Problem		
					CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A1502	Positioning Problem		
							F23	Unexpected Medical Intervention	A1502	Positioning Problem		
					CT233	Surgical	E2403	No Clinical Signs, Symptoms or Conditions	A1502	Positioning Problem		
							F23	Unexpected Medical Intervention	A1502	Positioning Problem		
					CT2406	Implantable	E2403	No Clinical Signs, Symptoms or Conditions	A1502	Positioning Problem		
							F23	Unexpected Medical Intervention	A1502	Positioning Problem		
					CT2407	Active implantable	E2403	No Clinical Signs, Symptoms or Conditions	A1502	Positioning Problem		
							F23	Unexpected Medical Intervention	A1502	Positioning Problem		
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A1502	Positioning Problem		
							F23	Unexpected Medical Intervention	A1502	Positioning Problem		
					CT2727	Active implant devices/ancillaries	E2403	No Clinical Signs, Symptoms or Conditions	A1502	Positioning Problem		
							F23	Unexpected Medical Intervention	A1502	Positioning Problem		
					CT2750	Implantable cardiac electrical stimulators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A1502	Positioning Problem		
							F23	Unexpected Medical Intervention	A1502	Positioning Problem		
					CT2818	Catheterization/Cannulation/Fluid path devices	E2403	No Clinical Signs, Symptoms or Conditions	A1502	Positioning Problem		
							F23	Unexpected Medical Intervention	A1502	Positioning Problem		
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A1502	Positioning Problem		
							F23	Unexpected Medical Intervention	A1502	Positioning Problem		
					CT2985	Active	E2403	No Clinical Signs, Symptoms or Conditions	A1502	Positioning Problem		
							F23	Unexpected Medical Intervention	A1502	Positioning Problem		
					CT3001	Implantable metal	E2403	No Clinical Signs, Symptoms or Conditions	A1502	Positioning Problem		
							F23	Unexpected Medical Intervention	A1502	Positioning Problem		
					CT3002	Implantable synthetic polymer	E2403	No Clinical Signs, Symptoms or Conditions	A1502	Positioning Problem		
							F23	Unexpected Medical Intervention	A1502	Positioning Problem		
					CT3006	Implantable inorganic material	E2403	No Clinical Signs, Symptoms or Conditions	A1502	Positioning Problem		
							F23	Unexpected Medical Intervention	A1502	Positioning Problem		

					CT314	Electrically powered	E2403	No Clinical Signs, Symptoms or Conditions	A1502	Positioning Problem		
							F23	Unexpected Medical Intervention	A1502	Positioning Problem		
					CT321	Long-term surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A1502	Positioning Problem		
							F23	Unexpected Medical Intervention	A1502	Positioning Problem		
					CT981	Single use	E2403	No Clinical Signs, Symptoms or Conditions	A1502	Positioning Problem		
							F23	Unexpected Medical Intervention	A1502	Positioning Problem		
					CT983	Surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A1502	Positioning Problem		
							F23	Unexpected Medical Intervention	A1502	Positioning Problem		
13610	2023	All other reportable incidents	17882	Physiologic-monitoring defibrillation system	CT1684	Cardiac pacemakers and associated devices			A071204	No Pacing		
13611	2023	All other reportable incidents	47265	Dual-chamber implantable pacemaker, rate-responsive	CT141	Electrodes/Leads and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A0908	Incorrect, Inadequate or Imprecise Result or Readings	10/07/2023	10/07/2023
							F08	Hospitalization or Prolonged Hospitalization	A0908	Incorrect, Inadequate or Imprecise Result or Readings	10/07/2023	10/07/2023
							F1905	Device Revision or Replacement	A0908	Incorrect, Inadequate or Imprecise Result or Readings	10/07/2023	10/07/2023
					CT145	Cardiovascular devices	E2403	No Clinical Signs, Symptoms or Conditions	A0908	Incorrect, Inadequate or Imprecise Result or Readings	10/07/2023	10/07/2023
							F08	Hospitalization or Prolonged Hospitalization	A0908	Incorrect, Inadequate or Imprecise Result or Readings	10/07/2023	10/07/2023
							F1905	Device Revision or Replacement	A0908	Incorrect, Inadequate or Imprecise Result or Readings	10/07/2023	10/07/2023
					CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A0908	Incorrect, Inadequate or Imprecise Result or Readings	10/07/2023	10/07/2023
							F08	Hospitalization or Prolonged Hospitalization	A0908	Incorrect, Inadequate or Imprecise Result or Readings	10/07/2023	10/07/2023
							F1905	Device Revision or Replacement	A0908	Incorrect, Inadequate or Imprecise Result or Readings	10/07/2023	10/07/2023
					CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A0908	Incorrect, Inadequate or Imprecise Result or Readings	10/07/2023	10/07/2023
							F08	Hospitalization or Prolonged Hospitalization	A0908	Incorrect, Inadequate or Imprecise Result or Readings	10/07/2023	10/07/2023
							F1905	Device Revision or Replacement	A0908	Incorrect, Inadequate or Imprecise Result or Readings	10/07/2023	10/07/2023
					CT233	Surgical	E2403	No Clinical Signs, Symptoms or Conditions	A0908	Incorrect, Inadequate or Imprecise Result or Readings	10/07/2023	10/07/2023
							F08	Hospitalization or Prolonged Hospitalization	A0908	Incorrect, Inadequate or Imprecise Result or Readings	10/07/2023	10/07/2023
							F1905	Device Revision or Replacement	A0908	Incorrect, Inadequate or Imprecise Result or Readings	10/07/2023	10/07/2023
					CT2406	Implantable	E2403	No Clinical Signs, Symptoms or Conditions	A0908	Incorrect, Inadequate or Imprecise Result or Readings	10/07/2023	10/07/2023
							F08	Hospitalization or Prolonged Hospitalization	A0908	Incorrect, Inadequate or Imprecise Result or Readings	10/07/2023	10/07/2023
							F1905	Device Revision or Replacement	A0908	Incorrect, Inadequate or Imprecise Result or Readings	10/07/2023	10/07/2023
					CT2407	Active implantable	E2403	No Clinical Signs, Symptoms or Conditions	A0908	Incorrect, Inadequate or Imprecise Result or Readings	10/07/2023	10/07/2023
							F08	Hospitalization or Prolonged Hospitalization	A0908	Incorrect, Inadequate or Imprecise Result or Readings	10/07/2023	10/07/2023
							F1905	Device Revision or Replacement	A0908	Incorrect, Inadequate or Imprecise Result or Readings	10/07/2023	10/07/2023
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A0908	Incorrect, Inadequate or Imprecise Result or Readings	10/07/2023	10/07/2023
							F08	Hospitalization or Prolonged Hospitalization	A0908	Incorrect, Inadequate or Imprecise Result or Readings	10/07/2023	10/07/2023
							F1905	Device Revision or Replacement	A0908	Incorrect, Inadequate or Imprecise Result or Readings	10/07/2023	10/07/2023
					CT2727	Active implant devices/ancillaries	E2403	No Clinical Signs, Symptoms or Conditions	A0908	Incorrect, Inadequate or Imprecise Result or Readings	10/07/2023	10/07/2023
							F08	Hospitalization or Prolonged Hospitalization	A0908	Incorrect, Inadequate or Imprecise Result or Readings	10/07/2023	10/07/2023
							F1905	Device Revision or Replacement	A0908	Incorrect, Inadequate or Imprecise Result or Readings	10/07/2023	10/07/2023
					CT2750	Implantable cardiac electrical stimulators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A0908	Incorrect, Inadequate or Imprecise Result or Readings	10/07/2023	10/07/2023
							F08	Hospitalization or Prolonged Hospitalization	A0908	Incorrect, Inadequate or Imprecise Result or Readings	10/07/2023	10/07/2023
							F1905	Device Revision or Replacement	A0908	Incorrect, Inadequate or Imprecise Result or Readings	10/07/2023	10/07/2023
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A0908	Incorrect, Inadequate or Imprecise Result or Readings	10/07/2023	10/07/2023
							F08	Hospitalization or Prolonged Hospitalization	A0908	Incorrect, Inadequate or Imprecise Result or Readings	10/07/2023	10/07/2023
							F1905	Device Revision or Replacement	A0908	Incorrect, Inadequate or Imprecise Result or Readings	10/07/2023	10/07/2023
					CT2985	Active	E2403	No Clinical Signs, Symptoms or Conditions	A0908	Incorrect, Inadequate or Imprecise Result or Readings	10/07/2023	10/07/2023
							F08	Hospitalization or Prolonged Hospitalization	A0908	Incorrect, Inadequate or Imprecise Result or Readings	10/07/2023	10/07/2023
							F1905	Device Revision or Replacement	A0908	Incorrect, Inadequate or Imprecise Result or Readings	10/07/2023	10/07/2023
					CT3001	Implantable metal	E2403	No Clinical Signs, Symptoms or Conditions	A0908	Incorrect, Inadequate or Imprecise Result or Readings	10/07/2023	10/07/2023
							F08	Hospitalization or Prolonged Hospitalization	A0908	Incorrect, Inadequate or Imprecise Result or Readings	10/07/2023	10/07/2023
							F1905	Device Revision or Replacement	A0908	Incorrect, Inadequate or Imprecise Result or Readings	10/07/2023	10/07/2023
					CT3002	Implantable synthetic polymer	E2403	No Clinical Signs, Symptoms or Conditions	A0908	Incorrect, Inadequate or Imprecise Result or Readings	10/07/2023	10/07/2023
							F08	Hospitalization or Prolonged Hospitalization	A0908	Incorrect, Inadequate or Imprecise Result or Readings	10/07/2023	10/07/2023
							F1905	Device Revision or Replacement	A0908	Incorrect, Inadequate or Imprecise Result or Readings	10/07/2023	10/07/2023
					CT3006	Implantable inorganic material	E2403	No Clinical Signs, Symptoms or Conditions	A0908	Incorrect, Inadequate or Imprecise Result or Readings	10/07/2023	10/07/2023
							F08	Hospitalization or Prolonged Hospitalization	A0908	Incorrect, Inadequate or Imprecise Result or Readings	10/07/2023	10/07/2023

							F1905	Device Revision or Replacement	A0908	Incorrect, Inadequate or Imprecise Result or Readings	10/07/2023	10/07/2023
					CT314	Electrically powered	E2403	No Clinical Signs, Symptoms or Conditions	A0908	Incorrect, Inadequate or Imprecise Result or Readings	10/07/2023	10/07/2023
							F08	Hospitalization or Prolonged Hospitalization	A0908	Incorrect, Inadequate or Imprecise Result or Readings	10/07/2023	10/07/2023
							F1905	Device Revision or Replacement	A0908	Incorrect, Inadequate or Imprecise Result or Readings	10/07/2023	10/07/2023
					CT321	Long-term surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A0908	Incorrect, Inadequate or Imprecise Result or Readings	10/07/2023	10/07/2023
							F08	Hospitalization or Prolonged Hospitalization	A0908	Incorrect, Inadequate or Imprecise Result or Readings	10/07/2023	10/07/2023
							F1905	Device Revision or Replacement	A0908	Incorrect, Inadequate or Imprecise Result or Readings	10/07/2023	10/07/2023
					CT449	Generators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A0908	Incorrect, Inadequate or Imprecise Result or Readings	10/07/2023	10/07/2023
							F08	Hospitalization or Prolonged Hospitalization	A0908	Incorrect, Inadequate or Imprecise Result or Readings	10/07/2023	10/07/2023
							F1905	Device Revision or Replacement	A0908	Incorrect, Inadequate or Imprecise Result or Readings	10/07/2023	10/07/2023
					CT981	Single use	E2403	No Clinical Signs, Symptoms or Conditions	A0908	Incorrect, Inadequate or Imprecise Result or Readings	10/07/2023	10/07/2023
							F08	Hospitalization or Prolonged Hospitalization	A0908	Incorrect, Inadequate or Imprecise Result or Readings	10/07/2023	10/07/2023
							F1905	Device Revision or Replacement	A0908	Incorrect, Inadequate or Imprecise Result or Readings	10/07/2023	10/07/2023
					CT983	Surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A0908	Incorrect, Inadequate or Imprecise Result or Readings	10/07/2023	10/07/2023
							F08	Hospitalization or Prolonged Hospitalization	A0908	Incorrect, Inadequate or Imprecise Result or Readings	10/07/2023	10/07/2023
							F1905	Device Revision or Replacement	A0908	Incorrect, Inadequate or Imprecise Result or Readings	10/07/2023	10/07/2023
13612	2023		47265	Dual-chamber implantable pacemaker, rate-responsive	CT141	Electrodes/Leads and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem		
							F1901	Additional Surgery	A24	No Apparent Device Problem		
					CT145	Cardiovascular devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem		
							F1901	Additional Surgery	A24	No Apparent Device Problem		
					CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem		
							F1901	Additional Surgery	A24	No Apparent Device Problem		
					CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem		
							F1901	Additional Surgery	A24	No Apparent Device Problem		
					CT233	Surgical	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem		
							F1901	Additional Surgery	A24	No Apparent Device Problem		
					CT2406	Implantable	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem		
							F1901	Additional Surgery	A24	No Apparent Device Problem		
					CT2407	Active implantable	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem		
							F1901	Additional Surgery	A24	No Apparent Device Problem		
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem		
							F1901	Additional Surgery	A24	No Apparent Device Problem		
					CT2727	Active implant devices/ancillaries	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem		
							F1901	Additional Surgery	A24	No Apparent Device Problem		
					CT2750	Implantable cardiac electrical stimulators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem		
							F1901	Additional Surgery	A24	No Apparent Device Problem		
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem		
							F1901	Additional Surgery	A24	No Apparent Device Problem		
					CT2985	Active	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem		
							F1901	Additional Surgery	A24	No Apparent Device Problem		
					CT3001	Implantable metal	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem		
							F1901	Additional Surgery	A24	No Apparent Device Problem		
					CT3002	Implantable synthetic polymer	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem		
							F1901	Additional Surgery	A24	No Apparent Device Problem		
					CT3006	Implantable inorganic material	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem		
							F1901	Additional Surgery	A24	No Apparent Device Problem		
					CT314	Electrically powered	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem		
							F1901	Additional Surgery	A24	No Apparent Device Problem		
					CT321	Long-term surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem		
							F1901	Additional Surgery	A24	No Apparent Device Problem		
					CT449	Generators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem		
							F1901	Additional Surgery	A24	No Apparent Device Problem		
					CT981	Single use	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem		
							F1901	Additional Surgery	A24	No Apparent Device Problem		
					CT983	Surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem		
							F1901	Additional Surgery	A24	No Apparent Device Problem		
13613	2023	All other reportable incidents	47205	Cardiac pulse generator programmer	CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A090805	Incorrect Measurement		
							F23	Unexpected Medical Intervention	A090805	Incorrect Measurement		
13614	2023	All other reportable incidents	47205	Cardiac pulse generator programmer	CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A090805	Incorrect Measurement		
							F23	Unexpected Medical Intervention	A090805	Incorrect Measurement		
13615	2023		47265	Dual-chamber implantable pacemaker, rate-responsive	CT141	Electrodes/Leads and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem		
							F1901	Additional Surgery	A24	No Apparent Device Problem		
					CT145	Cardiovascular devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem		
							F1901	Additional Surgery	A24	No Apparent Device Problem		
					CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem		
							F1901	Additional Surgery	A24	No Apparent Device Problem		
					CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem		
							F1901	Additional Surgery	A24	No Apparent Device Problem		
					CT233	Surgical	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem		
							F1901	Additional Surgery	A24	No Apparent Device Problem		
					CT2406	Implantable	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem		
							F1901	Additional Surgery	A24	No Apparent Device Problem		
					CT2407	Active implantable	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem		
							F1901	Additional Surgery	A24	No Apparent Device Problem		
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem		
							F1901	Additional Surgery	A24	No Apparent Device Problem		

					CT2727	Active implant devices/ancillaries	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem		
							F1901	Additional Surgery	A24	No Apparent Device Problem		
					CT2750	Implantable cardiac electrical stimulators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem		
							F1901	Additional Surgery	A24	No Apparent Device Problem		
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem		
							F1901	Additional Surgery	A24	No Apparent Device Problem		
					CT2985	Active	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem		
							F1901	Additional Surgery	A24	No Apparent Device Problem		
					CT3001	Implantable metal	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem		
							F1901	Additional Surgery	A24	No Apparent Device Problem		
					CT3002	Implantable synthetic polymer	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem		
							F1901	Additional Surgery	A24	No Apparent Device Problem		
					CT3006	Implantable inorganic material	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem		
							F1901	Additional Surgery	A24	No Apparent Device Problem		
					CT314	Electrically powered	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem		
							F1901	Additional Surgery	A24	No Apparent Device Problem		
					CT321	Long-term surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem		
							F1901	Additional Surgery	A24	No Apparent Device Problem		
					CT449	Generators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem		
							F1901	Additional Surgery	A24	No Apparent Device Problem		
					CT981	Single use	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem		
							F1901	Additional Surgery	A24	No Apparent Device Problem		
					CT983	Surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem		
							F1901	Additional Surgery	A24	No Apparent Device Problem		
13616	2023		47265	Dual-chamber implantable pacemaker, rate-responsive	CT141	Electrodes/Leads and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem		
							F1901	Additional Surgery	A24	No Apparent Device Problem		
					CT145	Cardiovascular devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem		
							F1901	Additional Surgery	A24	No Apparent Device Problem		
					CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem		
							F1901	Additional Surgery	A24	No Apparent Device Problem		
					CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem		
							F1901	Additional Surgery	A24	No Apparent Device Problem		
					CT233	Surgical	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem		
							F1901	Additional Surgery	A24	No Apparent Device Problem		
					CT2406	Implantable	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem		
							F1901	Additional Surgery	A24	No Apparent Device Problem		
					CT2407	Active implantable	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem		
							F1901	Additional Surgery	A24	No Apparent Device Problem		
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem		
							F1901	Additional Surgery	A24	No Apparent Device Problem		
					CT2727	Active implant devices/ancillaries	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem		
							F1901	Additional Surgery	A24	No Apparent Device Problem		
					CT2750	Implantable cardiac electrical stimulators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem		
							F1901	Additional Surgery	A24	No Apparent Device Problem		
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem		
							F1901	Additional Surgery	A24	No Apparent Device Problem		
					CT2985	Active	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem		
							F1901	Additional Surgery	A24	No Apparent Device Problem		
					CT3001	Implantable metal	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem		
							F1901	Additional Surgery	A24	No Apparent Device Problem		
					CT3002	Implantable synthetic polymer	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem		
							F1901	Additional Surgery	A24	No Apparent Device Problem		
					CT3006	Implantable inorganic material	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem		
							F1901	Additional Surgery	A24	No Apparent Device Problem		
					CT314	Electrically powered	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem		
							F1901	Additional Surgery	A24	No Apparent Device Problem		
					CT321	Long-term surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem		
							F1901	Additional Surgery	A24	No Apparent Device Problem		
					CT449	Generators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem		
							F1901	Additional Surgery	A24	No Apparent Device Problem		
					CT981	Single use	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem		
							F1901	Additional Surgery	A24	No Apparent Device Problem		
					CT983	Surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem		
							F1901	Additional Surgery	A24	No Apparent Device Problem		
13617	2023	All other reportable incidents	47205	Cardiac pulse generator programmer	CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A090805	Incorrect Measurement		
							F23	Unexpected Medical Intervention	A090805	Incorrect Measurement		
13618	2023	All other reportable incidents	47205	Cardiac pulse generator programmer	CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A090805	Incorrect Measurement		
							F23	Unexpected Medical Intervention	A090805	Incorrect Measurement		
13619	2023	All other reportable incidents	47205	Cardiac pulse generator programmer	CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A090805	Incorrect Measurement		
							F23	Unexpected Medical Intervention	A090805	Incorrect Measurement		
13620	2023		47265	Dual-chamber implantable pacemaker, rate-responsive	CT141	Electrodes/Leads and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	16/05/2023	16/05/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	16/05/2023	16/05/2023
					CT145	Cardiovascular devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	16/05/2023	16/05/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	16/05/2023	16/05/2023
					CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	16/05/2023	16/05/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	16/05/2023	16/05/2023
					CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	16/05/2023	16/05/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	16/05/2023	16/05/2023
					CT233	Surgical	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	16/05/2023	16/05/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	16/05/2023	16/05/2023
					CT2406	Implantable	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	16/05/2023	16/05/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	16/05/2023	16/05/2023
					CT2407	Active implantable	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	16/05/2023	16/05/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	16/05/2023	16/05/2023
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	16/05/2023	16/05/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	16/05/2023	16/05/2023
					CT2727	Active implant devices/ancillaries	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	16/05/2023	16/05/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	16/05/2023	16/05/2023

					CT2750	Implantable cardiac electrical stimulators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	16/05/2023	16/05/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	16/05/2023	16/05/2023
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	16/05/2023	16/05/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	16/05/2023	16/05/2023
					CT2985	Active	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	16/05/2023	16/05/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	16/05/2023	16/05/2023
					CT3001	Implantable metal	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	16/05/2023	16/05/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	16/05/2023	16/05/2023
					CT3002	Implantable synthetic polymer	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	16/05/2023	16/05/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	16/05/2023	16/05/2023
					CT3006	Implantable inorganic material	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	16/05/2023	16/05/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	16/05/2023	16/05/2023
					CT314	Electrically powered	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	16/05/2023	16/05/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	16/05/2023	16/05/2023
					CT321	Long-term surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	16/05/2023	16/05/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	16/05/2023	16/05/2023
					CT449	Generators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	16/05/2023	16/05/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	16/05/2023	16/05/2023
					CT981	Single use	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	16/05/2023	16/05/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	16/05/2023	16/05/2023
					CT983	Surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	16/05/2023	16/05/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	16/05/2023	16/05/2023
13621	2023	All other reportable incidents	47265	Dual-chamber implantable pacemaker, rate-responsive	CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A070102	High Capture Threshold	07/07/2023	07/07/2023
							F08	Hospitalization or Prolonged Hospitalization	A070102	High Capture Threshold	07/07/2023	07/07/2023
							F1905	Device Revision or Replacement	A070102	High Capture Threshold	07/07/2023	07/07/2023
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A070102	High Capture Threshold	07/07/2023	07/07/2023
							F08	Hospitalization or Prolonged Hospitalization	A070102	High Capture Threshold	07/07/2023	07/07/2023
							F1905	Device Revision or Replacement	A070102	High Capture Threshold	07/07/2023	07/07/2023
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A070102	High Capture Threshold	07/07/2023	07/07/2023
							F08	Hospitalization or Prolonged Hospitalization	A070102	High Capture Threshold	07/07/2023	07/07/2023
							F1905	Device Revision or Replacement	A070102	High Capture Threshold	07/07/2023	07/07/2023
13622	2023	All other reportable incidents	17882	Physiologic-monitoring defibrillation system	CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A071204	No Pacing		
							F27	Problem Identified During Non-Clinical Procedure	A071204	No Pacing		
13623	2023	All other reportable incidents	47205	Cardiac pulse generator programmer	CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A090805	Incorrect Measurement		
							F23	Unexpected Medical Intervention	A090805	Incorrect Measurement		
13624	2023		47265	Dual-chamber implantable pacemaker, rate-responsive	CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	14/04/2023	14/04/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	14/04/2023	14/04/2023
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	14/04/2023	14/04/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	14/04/2023	14/04/2023
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	14/04/2023	14/04/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	14/04/2023	14/04/2023
13625	2023	All other reportable incidents	17882	Physiologic-monitoring defibrillation system	CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A1603	Failure of Device to Self-Test		
							F27	Problem Identified During Non-Clinical Procedure	A1603	Failure of Device to Self-Test		
13626	2023	All other reportable incidents	35852	Single-chamber implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	04/07/2023	04/07/2023
									A0908	Incorrect, Inadequate or Imprecise Result or Readings	04/07/2023	04/07/2023
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	04/07/2023	04/07/2023
									A0908	Incorrect, Inadequate or Imprecise Result or Readings	04/07/2023	04/07/2023
							F1903	Device Explantation	A070504	Premature Discharge of Battery	04/07/2023	04/07/2023
									A0908	Incorrect, Inadequate or Imprecise Result or Readings	04/07/2023	04/07/2023
13627	2023	All other reportable incidents	35223	Endocardial/interventricular septal pacing lead	CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070901	Decreased Sensitivity		
									A070909	Oversensing		
							F26	No Health Consequences or Impact	A070901	Decreased Sensitivity		
									A070909	Oversensing		
13628	2023	All other reportable incidents	47267	Single-chamber implantable pacemaker, rate-responsive	CT141	Electrodes/Leads and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A0203	Defective Device		
							F26	No Health Consequences or Impact	A0203	Defective Device		
					CT145	Cardiovascular devices	E2403	No Clinical Signs, Symptoms or Conditions	A0203	Defective Device		
							F26	No Health Consequences or Impact	A0203	Defective Device		
					CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A0203	Defective Device		
							F26	No Health Consequences or Impact	A0203	Defective Device		
					CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A0203	Defective Device		
							F26	No Health Consequences or Impact	A0203	Defective Device		
					CT233	Surgical	E2403	No Clinical Signs, Symptoms or Conditions	A0203	Defective Device		
							F26	No Health Consequences or Impact	A0203	Defective Device		
					CT2406	Implantable	E2403	No Clinical Signs, Symptoms or Conditions	A0203	Defective Device		
							F26	No Health Consequences or Impact	A0203	Defective Device		
					CT2407	Active implantable	E2403	No Clinical Signs, Symptoms or Conditions	A0203	Defective Device		
							F26	No Health Consequences or Impact	A0203	Defective Device		
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A0203	Defective Device		
							F26	No Health Consequences or Impact	A0203	Defective Device		
					CT2727	Active implant devices/ancillaries	E2403	No Clinical Signs, Symptoms or Conditions	A0203	Defective Device		
							F26	No Health Consequences or Impact	A0203	Defective Device		
					CT2750	Implantable cardiac electrical stimulators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A0203	Defective Device		
							F26	No Health Consequences or Impact	A0203	Defective Device		
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A0203	Defective Device		
							F26	No Health Consequences or Impact	A0203	Defective Device		
					CT2985	Active	E2403	No Clinical Signs, Symptoms or Conditions	A0203	Defective Device		
							F26	No Health Consequences or Impact	A0203	Defective Device		
					CT3001	Implantable metal	E2403	No Clinical Signs, Symptoms or Conditions	A0203	Defective Device		
							F26	No Health Consequences or Impact	A0203	Defective Device		
					CT3002	Implantable synthetic polymer	E2403	No Clinical Signs, Symptoms or Conditions	A0203	Defective Device		
							F26	No Health Consequences or Impact	A0203	Defective Device		
					CT3006	Implantable inorganic material	E2403	No Clinical Signs, Symptoms or Conditions	A0203	Defective Device		
							F26	No Health Consequences or Impact	A0203	Defective Device		
					CT314	Electrically powered	E2403	No Clinical Signs, Symptoms or Conditions	A0203	Defective Device		
							F26	No Health Consequences or Impact	A0203	Defective Device		

					CT321	Long-term surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A0203	Defective Device		
							F26	No Health Consequences or Impact	A0203	Defective Device		
					CT449	Generators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A0203	Defective Device		
							F26	No Health Consequences or Impact	A0203	Defective Device		
					CT981	Single use	E2403	No Clinical Signs, Symptoms or Conditions	A0203	Defective Device		
							F26	No Health Consequences or Impact	A0203	Defective Device		
					CT983	Surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A0203	Defective Device		
							F26	No Health Consequences or Impact	A0203	Defective Device		
13629	2023	Unanticipated serious deterioration in state of health	35223	Endocardial/interventricular septal pacing lead	CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A0722	Impedance Problem		
							F1905	Device Revision or Replacement	A0722	Impedance Problem		
13630	2023	All other reportable incidents	47265	Dual-chamber implantable pacemaker, rate-responsive	CT141	Electrodes/Leads and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	01/06/2023	01/06/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	01/06/2023	01/06/2023
					CT145	Cardiovascular devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	01/06/2023	01/06/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	01/06/2023	01/06/2023
					CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	01/06/2023	01/06/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	01/06/2023	01/06/2023
					CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	01/06/2023	01/06/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	01/06/2023	01/06/2023
					CT233	Surgical	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	01/06/2023	01/06/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	01/06/2023	01/06/2023
					CT2406	Implantable	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	01/06/2023	01/06/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	01/06/2023	01/06/2023
					CT2407	Active implantable	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	01/06/2023	01/06/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	01/06/2023	01/06/2023
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	01/06/2023	01/06/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	01/06/2023	01/06/2023
					CT2727	Active implant devices/ancillaries	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	01/06/2023	01/06/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	01/06/2023	01/06/2023
					CT2750	Implantable cardiac electrical stimulators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	01/06/2023	01/06/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	01/06/2023	01/06/2023
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	01/06/2023	01/06/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	01/06/2023	01/06/2023
					CT2985	Active	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	01/06/2023	01/06/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	01/06/2023	01/06/2023
					CT3001	Implantable metal	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	01/06/2023	01/06/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	01/06/2023	01/06/2023
					CT3002	Implantable synthetic polymer	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	01/06/2023	01/06/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	01/06/2023	01/06/2023
					CT3006	Implantable inorganic material	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	01/06/2023	01/06/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	01/06/2023	01/06/2023
					CT314	Electrically powered	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	01/06/2023	01/06/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	01/06/2023	01/06/2023
					CT321	Long-term surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	01/06/2023	01/06/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	01/06/2023	01/06/2023
					CT449	Generators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	01/06/2023	01/06/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	01/06/2023	01/06/2023
					CT981	Single use	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	01/06/2023	01/06/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	01/06/2023	01/06/2023
					CT983	Surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	01/06/2023	01/06/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	01/06/2023	01/06/2023
13631	2023	All other reportable incidents	35223	Endocardial/interventricular septal pacing lead	CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070909	Oversensing		
							F1908	Prolonged Surgery	A070909	Oversensing		
13632	2023	All other reportable incidents	47265	Dual-chamber implantable pacemaker, rate-responsive	CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	14/04/2023	14/04/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	14/04/2023	14/04/2023
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	14/04/2023	14/04/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	14/04/2023	14/04/2023
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	14/04/2023	14/04/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	14/04/2023	14/04/2023
13633	2023	All other reportable incidents	35852	Single-chamber implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	13/03/2023	13/03/2023
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	13/03/2023	13/03/2023
							F12	Serious Injury/ Illness/ Impairment	A070504	Premature Discharge of Battery	13/03/2023	13/03/2023
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	13/03/2023	13/03/2023
13634	2023	All other reportable incidents	35852	Single-chamber implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A0713	Defibrillation/Stimulation Problem		
							F12	Serious Injury/ Illness/ Impairment	A0713	Defibrillation/Stimulation Problem		
							F23	Unexpected Medical Intervention	A0713	Defibrillation/Stimulation Problem		
13635	2023	All other reportable incidents	17882	Physiologic-monitoring defibrillation system	CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A0719	Unexpected Shutdown		
							F26	No Health Consequences or Impact	A0719	Unexpected Shutdown		
13636	2023	All other reportable incidents	36241	Pacemaker/defibrillator lead	CT1684	Cardiac pacemakers and associated devices	E0131	Speech Disorder	A010402	Migration		
									A04	Material Integrity Problem		
									A040101	Fracture		
									A070101	Failure to Capture		
									A090801	Signal Artifact/Noise		
							E013302	Ischemia Stroke	A010402	Migration		
									A04	Material Integrity Problem		
									A040101	Fracture		
									A070101	Failure to Capture		
									A090801	Signal Artifact/Noise		
							E013403	Paresthesia	A010402	Migration		
									A04	Material Integrity Problem		
									A040101	Fracture		
									A070101	Failure to Capture		
									A090801	Signal Artifact/Noise		
							E0137	Transient Ischemic Attack	A010402	Migration		
									A04	Material Integrity Problem		
									A040101	Fracture		
									A070101	Failure to Capture		
									A090801	Signal Artifact/Noise		
							E050304	Thromboembolism	A010402	Migration		

									A04	Material Integrity Problem		
									A040101	Fracture		
									A070101	Failure to Capture		
									A090801	Signal Artifact/Noise		
							E083903	Diplopia/Double Vision	A010402	Migration		
									A04	Material Integrity Problem		
									A040101	Fracture		
									A070101	Failure to Capture		
									A090801	Signal Artifact/Noise		
							F08	Hospitalization or Prolonged Hospitalization	A010402	Migration		
									A04	Material Integrity Problem		
									A040101	Fracture		
									A070101	Failure to Capture		
									A090801	Signal Artifact/Noise		
							F1203	Life Threatening Illness or Injury	A010402	Migration		
									A04	Material Integrity Problem		
									A040101	Fracture		
									A070101	Failure to Capture		
									A090801	Signal Artifact/Noise		
							F1204	Permanent Impairment	A010402	Migration		
									A04	Material Integrity Problem		
									A040101	Fracture		
									A070101	Failure to Capture		
									A090801	Signal Artifact/Noise		
							F1905	Device Revision or Replacement	A010402	Migration		
									A04	Material Integrity Problem		
									A040101	Fracture		
									A070101	Failure to Capture		
									A090801	Signal Artifact/Noise		
							F2203	Imaging Required	A010402	Migration		
									A04	Material Integrity Problem		
									A040101	Fracture		
									A070101	Failure to Capture		
									A090801	Signal Artifact/Noise		
							F2303	Medication Required	A010402	Migration		
									A04	Material Integrity Problem		
									A040101	Fracture		
									A070101	Failure to Capture		
									A090801	Signal Artifact/Noise		
13637	2023	All other reportable incidents	47263	Cardiac resynchronization therapy implantable pacemaker	CT1684	Cardiac pacemakers and associated devices	E2103	Device Overstimulation of Tissue	A071209	Extracardiac Stimulation	17/04/2023	17/04/2023
									A0908	Incorrect, Inadequate or Imprecise Result or Readings	17/04/2023	17/04/2023
							F08	Hospitalization or Prolonged Hospitalization	A071209	Extracardiac Stimulation	17/04/2023	17/04/2023
									A0908	Incorrect, Inadequate or Imprecise Result or Readings	17/04/2023	17/04/2023
							F1903	Device Explantation	A071209	Extracardiac Stimulation	17/04/2023	17/04/2023
									A0908	Incorrect, Inadequate or Imprecise Result or Readings	17/04/2023	17/04/2023
					CT1685	Cardiac pacemakers	E2103	Device Overstimulation of Tissue	A071209	Extracardiac Stimulation	17/04/2023	17/04/2023
									A0908	Incorrect, Inadequate or Imprecise Result or Readings	17/04/2023	17/04/2023
							F08	Hospitalization or Prolonged Hospitalization	A071209	Extracardiac Stimulation	17/04/2023	17/04/2023
									A0908	Incorrect, Inadequate or Imprecise Result or Readings	17/04/2023	17/04/2023
							F1903	Device Explantation	A071209	Extracardiac Stimulation	17/04/2023	17/04/2023
									A0908	Incorrect, Inadequate or Imprecise Result or Readings	17/04/2023	17/04/2023
					CT257	Implantable cardiac pacemakers	E2103	Device Overstimulation of Tissue	A071209	Extracardiac Stimulation	17/04/2023	17/04/2023
									A0908	Incorrect, Inadequate or Imprecise Result or Readings	17/04/2023	17/04/2023
							F08	Hospitalization or Prolonged Hospitalization	A071209	Extracardiac Stimulation	17/04/2023	17/04/2023
									A0908	Incorrect, Inadequate or Imprecise Result or Readings	17/04/2023	17/04/2023
							F1903	Device Explantation	A071209	Extracardiac Stimulation	17/04/2023	17/04/2023
									A0908	Incorrect, Inadequate or Imprecise Result or Readings	17/04/2023	17/04/2023
					CT2866	Implantable cardiac electrical stimulators	E2103	Device Overstimulation of Tissue	A071209	Extracardiac Stimulation	17/04/2023	17/04/2023
									A0908	Incorrect, Inadequate or Imprecise Result or Readings	17/04/2023	17/04/2023
							F08	Hospitalization or Prolonged Hospitalization	A071209	Extracardiac Stimulation	17/04/2023	17/04/2023
									A0908	Incorrect, Inadequate or Imprecise Result or Readings	17/04/2023	17/04/2023
							F1903	Device Explantation	A071209	Extracardiac Stimulation	17/04/2023	17/04/2023
									A0908	Incorrect, Inadequate or Imprecise Result or Readings	17/04/2023	17/04/2023
13638	2023		47265	Dual-chamber implantable pacemaker, rate-responsive	CT141	Electrodes/Leads and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	16/06/2023	16/06/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	16/06/2023	16/06/2023
					CT145	Cardiovascular devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	16/06/2023	16/06/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	16/06/2023	16/06/2023
					CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	16/06/2023	16/06/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	16/06/2023	16/06/2023
					CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	16/06/2023	16/06/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	16/06/2023	16/06/2023
					CT233	Surgical	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	16/06/2023	16/06/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	16/06/2023	16/06/2023
					CT2406	Implantable	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	16/06/2023	16/06/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	16/06/2023	16/06/2023
					CT2407	Active implantable	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	16/06/2023	16/06/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	16/06/2023	16/06/2023
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	16/06/2023	16/06/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	16/06/2023	16/06/2023

					CT2727	Active implant devices/ancillaries	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	16/06/2023	16/06/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	16/06/2023	16/06/2023
					CT2750	Implantable cardiac electrical stimulators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	16/06/2023	16/06/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	16/06/2023	16/06/2023
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	16/06/2023	16/06/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	16/06/2023	16/06/2023
					CT2985	Active	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	16/06/2023	16/06/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	16/06/2023	16/06/2023
					CT3001	Implantable metal	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	16/06/2023	16/06/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	16/06/2023	16/06/2023
					CT3002	Implantable synthetic polymer	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	16/06/2023	16/06/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	16/06/2023	16/06/2023
					CT3006	Implantable inorganic material	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	16/06/2023	16/06/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	16/06/2023	16/06/2023
					CT314	Electrically powered	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	16/06/2023	16/06/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	16/06/2023	16/06/2023
					CT321	Long-term surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	16/06/2023	16/06/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	16/06/2023	16/06/2023
					CT449	Generators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	16/06/2023	16/06/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	16/06/2023	16/06/2023
					CT981	Single use	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	16/06/2023	16/06/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	16/06/2023	16/06/2023
					CT983	Surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	16/06/2023	16/06/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	16/06/2023	16/06/2023
13639	2023		47265	Dual-chamber implantable pacemaker, rate-responsive	CT141	Electrodes/Leads and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	15/06/2023	15/06/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	15/06/2023	15/06/2023
					CT145	Cardiovascular devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	15/06/2023	15/06/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	15/06/2023	15/06/2023
					CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	15/06/2023	15/06/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	15/06/2023	15/06/2023
					CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	15/06/2023	15/06/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	15/06/2023	15/06/2023
					CT233	Surgical	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	15/06/2023	15/06/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	15/06/2023	15/06/2023
					CT2406	Implantable	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	15/06/2023	15/06/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	15/06/2023	15/06/2023
					CT2407	Active implantable	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	15/06/2023	15/06/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	15/06/2023	15/06/2023
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	15/06/2023	15/06/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	15/06/2023	15/06/2023
					CT2727	Active implant devices/ancillaries	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	15/06/2023	15/06/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	15/06/2023	15/06/2023
					CT2750	Implantable cardiac electrical stimulators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	15/06/2023	15/06/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	15/06/2023	15/06/2023
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	15/06/2023	15/06/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	15/06/2023	15/06/2023
					CT2985	Active	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	15/06/2023	15/06/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	15/06/2023	15/06/2023
					CT3001	Implantable metal	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	15/06/2023	15/06/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	15/06/2023	15/06/2023
					CT3002	Implantable synthetic polymer	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	15/06/2023	15/06/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	15/06/2023	15/06/2023
					CT3006	Implantable inorganic material	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	15/06/2023	15/06/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	15/06/2023	15/06/2023
					CT314	Electrically powered	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	15/06/2023	15/06/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	15/06/2023	15/06/2023
					CT321	Long-term surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	15/06/2023	15/06/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	15/06/2023	15/06/2023
					CT449	Generators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	15/06/2023	15/06/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	15/06/2023	15/06/2023
					CT981	Single use	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	15/06/2023	15/06/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	15/06/2023	15/06/2023
					CT983	Surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	15/06/2023	15/06/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	15/06/2023	15/06/2023
13640	2023	All other reportable incidents	47265	Dual-chamber implantable pacemaker, rate-responsive	CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem		
							F1901	Additional Surgery	A24	No Apparent Device Problem		
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem		
							F1901	Additional Surgery	A24	No Apparent Device Problem		
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem		
							F1901	Additional Surgery	A24	No Apparent Device Problem		
13641	2023		47265	Dual-chamber implantable pacemaker, rate-responsive	CT141	Electrodes/Leads and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem		
							F1901	Additional Surgery	A24	No Apparent Device Problem		
					CT145	Cardiovascular devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem		
							F1901	Additional Surgery	A24	No Apparent Device Problem		
					CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem		
							F1901	Additional Surgery	A24	No Apparent Device Problem		
					CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem		
							F1901	Additional Surgery	A24	No Apparent Device Problem		
					CT233	Surgical	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem		
							F1901	Additional Surgery	A24	No Apparent Device Problem		
					CT2406	Implantable	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem		
							F1901	Additional Surgery	A24	No Apparent Device Problem		
					CT2407	Active implantable	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem		
							F1901	Additional Surgery	A24	No Apparent Device Problem		
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem		
							F1901	Additional Surgery	A24	No Apparent Device Problem		
					CT2727	Active implant devices/ancillaries	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem		
							F1901	Additional Surgery	A24	No Apparent Device Problem		

					CT2750	Implantable cardiac electrical stimulators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem		
							F1901	Additional Surgery	A24	No Apparent Device Problem		
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem		
							F1901	Additional Surgery	A24	No Apparent Device Problem		
					CT2985	Active	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem		
							F1901	Additional Surgery	A24	No Apparent Device Problem		
					CT3001	Implantable metal	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem		
							F1901	Additional Surgery	A24	No Apparent Device Problem		
					CT3002	Implantable synthetic polymer	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem		
							F1901	Additional Surgery	A24	No Apparent Device Problem		
					CT3006	Implantable inorganic material	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem		
							F1901	Additional Surgery	A24	No Apparent Device Problem		
					CT314	Electrically powered	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem		
							F1901	Additional Surgery	A24	No Apparent Device Problem		
					CT321	Long-term surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem		
							F1901	Additional Surgery	A24	No Apparent Device Problem		
					CT449	Generators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem		
							F1901	Additional Surgery	A24	No Apparent Device Problem		
					CT981	Single use	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem		
							F1901	Additional Surgery	A24	No Apparent Device Problem		
					CT983	Surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem		
							F1901	Additional Surgery	A24	No Apparent Device Problem		
13642	2023		47265	Dual-chamber implantable pacemaker, rate-responsive	CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem		
							F1901	Additional Surgery	A24	No Apparent Device Problem		
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem		
							F1901	Additional Surgery	A24	No Apparent Device Problem		
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem		
							F1901	Additional Surgery	A24	No Apparent Device Problem		
13643	2023		47265	Dual-chamber implantable pacemaker, rate-responsive	CT141	Electrodes/Leads and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem		
							F1901	Additional Surgery	A24	No Apparent Device Problem		
					CT145	Cardiovascular devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem		
							F1901	Additional Surgery	A24	No Apparent Device Problem		
					CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem		
							F1901	Additional Surgery	A24	No Apparent Device Problem		
					CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem		
							F1901	Additional Surgery	A24	No Apparent Device Problem		
					CT233	Surgical	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem		
							F1901	Additional Surgery	A24	No Apparent Device Problem		
					CT2406	Implantable	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem		
							F1901	Additional Surgery	A24	No Apparent Device Problem		
					CT2407	Active implantable	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem		
							F1901	Additional Surgery	A24	No Apparent Device Problem		
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem		
							F1901	Additional Surgery	A24	No Apparent Device Problem		
					CT2727	Active implant devices/ancillaries	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem		
							F1901	Additional Surgery	A24	No Apparent Device Problem		
					CT2750	Implantable cardiac electrical stimulators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem		
							F1901	Additional Surgery	A24	No Apparent Device Problem		
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem		
							F1901	Additional Surgery	A24	No Apparent Device Problem		
					CT2985	Active	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem		
							F1901	Additional Surgery	A24	No Apparent Device Problem		
					CT3001	Implantable metal	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem		
							F1901	Additional Surgery	A24	No Apparent Device Problem		
					CT3002	Implantable synthetic polymer	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem		
							F1901	Additional Surgery	A24	No Apparent Device Problem		
					CT3006	Implantable inorganic material	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem		
							F1901	Additional Surgery	A24	No Apparent Device Problem		
					CT314	Electrically powered	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem		
							F1901	Additional Surgery	A24	No Apparent Device Problem		
					CT321	Long-term surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem		
							F1901	Additional Surgery	A24	No Apparent Device Problem		
					CT449	Generators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem		
							F1901	Additional Surgery	A24	No Apparent Device Problem		
					CT981	Single use	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem		
							F1901	Additional Surgery	A24	No Apparent Device Problem		
					CT983	Surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem		
							F1901	Additional Surgery	A24	No Apparent Device Problem		
13644	2023		47265	Dual-chamber implantable pacemaker, rate-responsive	CT141	Electrodes/Leads and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem		
							F1901	Additional Surgery	A24	No Apparent Device Problem		
					CT145	Cardiovascular devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem		
							F1901	Additional Surgery	A24	No Apparent Device Problem		
					CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem		
							F1901	Additional Surgery	A24	No Apparent Device Problem		
					CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem		
							F1901	Additional Surgery	A24	No Apparent Device Problem		
					CT233	Surgical	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem		
							F1901	Additional Surgery	A24	No Apparent Device Problem		
					CT2406	Implantable	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem		
							F1901	Additional Surgery	A24	No Apparent Device Problem		
					CT2407	Active implantable	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem		
							F1901	Additional Surgery	A24	No Apparent Device Problem		
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem		
							F1901	Additional Surgery	A24	No Apparent Device Problem		
					CT2727	Active implant devices/ancillaries	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem		
							F1901	Additional Surgery	A24	No Apparent Device Problem		
					CT2750	Implantable cardiac electrical stimulators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem		
							F1901	Additional Surgery	A24	No Apparent Device Problem		

					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem		
							F1901	Additional Surgery	A24	No Apparent Device Problem		
					CT2985	Active	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem		
							F1901	Additional Surgery	A24	No Apparent Device Problem		
					CT3001	Implantable metal	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem		
							F1901	Additional Surgery	A24	No Apparent Device Problem		
					CT3002	Implantable synthetic polymer	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem		
							F1901	Additional Surgery	A24	No Apparent Device Problem		
					CT3006	Implantable inorganic material	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem		
							F1901	Additional Surgery	A24	No Apparent Device Problem		
					CT314	Electrically powered	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem		
							F1901	Additional Surgery	A24	No Apparent Device Problem		
					CT321	Long-term surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem		
							F1901	Additional Surgery	A24	No Apparent Device Problem		
					CT449	Generators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem		
							F1901	Additional Surgery	A24	No Apparent Device Problem		
					CT981	Single use	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem		
							F1901	Additional Surgery	A24	No Apparent Device Problem		
					CT983	Surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem		
							F1901	Additional Surgery	A24	No Apparent Device Problem		
13645	2023		47265	Dual-chamber implantable pacemaker, rate-responsive	CT141	Electrodes/Leads and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem		
							F1901	Additional Surgery	A24	No Apparent Device Problem		
					CT145	Cardiovascular devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem		
							F1901	Additional Surgery	A24	No Apparent Device Problem		
					CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem		
							F1901	Additional Surgery	A24	No Apparent Device Problem		
					CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem		
							F1901	Additional Surgery	A24	No Apparent Device Problem		
					CT233	Surgical	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem		
							F1901	Additional Surgery	A24	No Apparent Device Problem		
					CT2406	Implantable	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem		
							F1901	Additional Surgery	A24	No Apparent Device Problem		
					CT2407	Active implantable	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem		
							F1901	Additional Surgery	A24	No Apparent Device Problem		
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem		
							F1901	Additional Surgery	A24	No Apparent Device Problem		
					CT2727	Active implant devices/ancillaries	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem		
							F1901	Additional Surgery	A24	No Apparent Device Problem		
					CT2750	Implantable cardiac electrical stimulators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem		
							F1901	Additional Surgery	A24	No Apparent Device Problem		
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem		
							F1901	Additional Surgery	A24	No Apparent Device Problem		
					CT2985	Active	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem		
							F1901	Additional Surgery	A24	No Apparent Device Problem		
					CT3001	Implantable metal	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem		
							F1901	Additional Surgery	A24	No Apparent Device Problem		
					CT3002	Implantable synthetic polymer	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem		
							F1901	Additional Surgery	A24	No Apparent Device Problem		
					CT3006	Implantable inorganic material	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem		
							F1901	Additional Surgery	A24	No Apparent Device Problem		
					CT314	Electrically powered	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem		
							F1901	Additional Surgery	A24	No Apparent Device Problem		
					CT321	Long-term surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem		
							F1901	Additional Surgery	A24	No Apparent Device Problem		
					CT449	Generators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem		
							F1901	Additional Surgery	A24	No Apparent Device Problem		
					CT981	Single use	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem		
							F1901	Additional Surgery	A24	No Apparent Device Problem		
					CT983	Surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem		
							F1901	Additional Surgery	A24	No Apparent Device Problem		
13646	2023		47265	Dual-chamber implantable pacemaker, rate-responsive	CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem		
							F1901	Additional Surgery	A24	No Apparent Device Problem		
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem		
							F1901	Additional Surgery	A24	No Apparent Device Problem		
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem		
							F1901	Additional Surgery	A24	No Apparent Device Problem		
13647	2023		47265	Dual-chamber implantable pacemaker, rate-responsive	CT141	Electrodes/Leads and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem		
							F1901	Additional Surgery	A24	No Apparent Device Problem		
					CT145	Cardiovascular devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem		
							F1901	Additional Surgery	A24	No Apparent Device Problem		
					CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem		
							F1901	Additional Surgery	A24	No Apparent Device Problem		
					CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem		
							F1901	Additional Surgery	A24	No Apparent Device Problem		
					CT233	Surgical	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem		
							F1901	Additional Surgery	A24	No Apparent Device Problem		
					CT2406	Implantable	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem		
							F1901	Additional Surgery	A24	No Apparent Device Problem		
					CT2407	Active implantable	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem		
							F1901	Additional Surgery	A24	No Apparent Device Problem		
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem		
							F1901	Additional Surgery	A24	No Apparent Device Problem		
					CT2727	Active implant devices/ancillaries	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem		
							F1901	Additional Surgery	A24	No Apparent Device Problem		
					CT2750	Implantable cardiac electrical stimulators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem		
							F1901	Additional Surgery	A24	No Apparent Device Problem		
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem		
							F1901	Additional Surgery	A24	No Apparent Device Problem		

					CT2985	Active	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem		
							F1901	Additional Surgery	A24	No Apparent Device Problem		
					CT3001	Implantable metal	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem		
							F1901	Additional Surgery	A24	No Apparent Device Problem		
					CT3002	Implantable synthetic polymer	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem		
							F1901	Additional Surgery	A24	No Apparent Device Problem		
					CT3006	Implantable inorganic material	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem		
							F1901	Additional Surgery	A24	No Apparent Device Problem		
					CT314	Electrically powered	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem		
							F1901	Additional Surgery	A24	No Apparent Device Problem		
					CT321	Long-term surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem		
							F1901	Additional Surgery	A24	No Apparent Device Problem		
					CT449	Generators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem		
							F1901	Additional Surgery	A24	No Apparent Device Problem		
					CT981	Single use	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem		
							F1901	Additional Surgery	A24	No Apparent Device Problem		
					CT983	Surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem		
							F1901	Additional Surgery	A24	No Apparent Device Problem		
13648	2023		47265	Dual-chamber implantable pacemaker, rate-responsive	CT141	Electrodes/Leads and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem		
							F1901	Additional Surgery	A24	No Apparent Device Problem		
					CT145	Cardiovascular devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem		
							F1901	Additional Surgery	A24	No Apparent Device Problem		
					CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem		
							F1901	Additional Surgery	A24	No Apparent Device Problem		
					CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem		
							F1901	Additional Surgery	A24	No Apparent Device Problem		
					CT233	Surgical	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem		
							F1901	Additional Surgery	A24	No Apparent Device Problem		
					CT2406	Implantable	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem		
							F1901	Additional Surgery	A24	No Apparent Device Problem		
					CT2407	Active implantable	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem		
							F1901	Additional Surgery	A24	No Apparent Device Problem		
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem		
							F1901	Additional Surgery	A24	No Apparent Device Problem		
					CT2727	Active implant devices/ancillaries	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem		
							F1901	Additional Surgery	A24	No Apparent Device Problem		
					CT2750	Implantable cardiac electrical stimulators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem		
							F1901	Additional Surgery	A24	No Apparent Device Problem		
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem		
							F1901	Additional Surgery	A24	No Apparent Device Problem		
					CT2985	Active	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem		
							F1901	Additional Surgery	A24	No Apparent Device Problem		
					CT3001	Implantable metal	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem		
							F1901	Additional Surgery	A24	No Apparent Device Problem		
					CT3002	Implantable synthetic polymer	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem		
							F1901	Additional Surgery	A24	No Apparent Device Problem		
					CT3006	Implantable inorganic material	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem		
							F1901	Additional Surgery	A24	No Apparent Device Problem		
					CT314	Electrically powered	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem		
							F1901	Additional Surgery	A24	No Apparent Device Problem		
					CT321	Long-term surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem		
							F1901	Additional Surgery	A24	No Apparent Device Problem		
					CT449	Generators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem		
							F1901	Additional Surgery	A24	No Apparent Device Problem		
					CT981	Single use	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem		
							F1901	Additional Surgery	A24	No Apparent Device Problem		
					CT983	Surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem		
							F1901	Additional Surgery	A24	No Apparent Device Problem		
13649	2023		47265	Dual-chamber implantable pacemaker, rate-responsive	CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem		
							F1901	Additional Surgery	A24	No Apparent Device Problem		
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem		
							F1901	Additional Surgery	A24	No Apparent Device Problem		
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem		
							F1901	Additional Surgery	A24	No Apparent Device Problem		
13650	2023	All other reportable incidents	47270	Cardiac resynchronization therapy implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E2104	Electric Shock	A0709	Device Sensing Problem		
									A070909	Oversensing		
									A071302	Inappropriate/Inadequate Shock/Stimulation		
							F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A0709	Device Sensing Problem		
									A070909	Oversensing		
									A071302	Inappropriate/Inadequate Shock/Stimulation		
13651	2023	All other reportable incidents	35852	Single-chamber implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E2104	Electric Shock	A0709	Device Sensing Problem		
									A070909	Oversensing		
									A071302	Inappropriate/Inadequate Shock/Stimulation		
							F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A0709	Device Sensing Problem		
									A070909	Oversensing		
									A071302	Inappropriate/Inadequate Shock/Stimulation		
13652	2023	All other reportable incidents	35223	Endocardial/interventricular septal pacing lead	CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A07	Electrical /Electronic Property Problem		
									A070909	Oversensing		
									A072202	Low Impedance		
							F26	No Health Consequences or Impact	A07	Electrical /Electronic Property Problem		
									A070909	Oversensing		
									A072202	Low Impedance		
13653	2023	All other reportable incidents	47265	Dual-chamber implantable pacemaker, rate-responsive	CT141	Electrodes/Leads and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A072201	High Impedance		
							F26	No Health Consequences or Impact	A072201	High Impedance		
					CT145	Cardiovascular devices	E2403	No Clinical Signs, Symptoms or Conditions	A072201	High Impedance		
							F26	No Health Consequences or Impact	A072201	High Impedance		
					CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A072201	High Impedance		
							F26	No Health Consequences or Impact	A072201	High Impedance		

					CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A072201	High Impedance		
							F26	No Health Consequences or Impact	A072201	High Impedance		
					CT233	Surgical	E2403	No Clinical Signs, Symptoms or Conditions	A072201	High Impedance		
							F26	No Health Consequences or Impact	A072201	High Impedance		
					CT2406	Implantable	E2403	No Clinical Signs, Symptoms or Conditions	A072201	High Impedance		
							F26	No Health Consequences or Impact	A072201	High Impedance		
					CT2407	Active implantable	E2403	No Clinical Signs, Symptoms or Conditions	A072201	High Impedance		
							F26	No Health Consequences or Impact	A072201	High Impedance		
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A072201	High Impedance		
							F26	No Health Consequences or Impact	A072201	High Impedance		
					CT2727	Active implant devices/ancillaries	E2403	No Clinical Signs, Symptoms or Conditions	A072201	High Impedance		
							F26	No Health Consequences or Impact	A072201	High Impedance		
					CT2750	Implantable cardiac electrical stimulators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A072201	High Impedance		
							F26	No Health Consequences or Impact	A072201	High Impedance		
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A072201	High Impedance		
							F26	No Health Consequences or Impact	A072201	High Impedance		
					CT2985	Active	E2403	No Clinical Signs, Symptoms or Conditions	A072201	High Impedance		
							F26	No Health Consequences or Impact	A072201	High Impedance		
					CT3001	Implantable metal	E2403	No Clinical Signs, Symptoms or Conditions	A072201	High Impedance		
							F26	No Health Consequences or Impact	A072201	High Impedance		
					CT3002	Implantable synthetic polymer	E2403	No Clinical Signs, Symptoms or Conditions	A072201	High Impedance		
							F26	No Health Consequences or Impact	A072201	High Impedance		
					CT3006	Implantable inorganic material	E2403	No Clinical Signs, Symptoms or Conditions	A072201	High Impedance		
							F26	No Health Consequences or Impact	A072201	High Impedance		
					CT314	Electrically powered	E2403	No Clinical Signs, Symptoms or Conditions	A072201	High Impedance		
							F26	No Health Consequences or Impact	A072201	High Impedance		
					CT321	Long-term surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A072201	High Impedance		
							F26	No Health Consequences or Impact	A072201	High Impedance		
					CT449	Generators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A072201	High Impedance		
							F26	No Health Consequences or Impact	A072201	High Impedance		
					CT981	Single use	E2403	No Clinical Signs, Symptoms or Conditions	A072201	High Impedance		
							F26	No Health Consequences or Impact	A072201	High Impedance		
					CT983	Surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A072201	High Impedance		
							F26	No Health Consequences or Impact	A072201	High Impedance		
13654	2023		47265	Dual-chamber implantable pacemaker, rate-responsive	CT141	Electrodes/Leads and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem		
							F1901	Additional Surgery	A24	No Apparent Device Problem		
					CT145	Cardiovascular devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem		
							F1901	Additional Surgery	A24	No Apparent Device Problem		
					CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem		
							F1901	Additional Surgery	A24	No Apparent Device Problem		
					CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem		
							F1901	Additional Surgery	A24	No Apparent Device Problem		
					CT233	Surgical	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem		
							F1901	Additional Surgery	A24	No Apparent Device Problem		
					CT2406	Implantable	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem		
							F1901	Additional Surgery	A24	No Apparent Device Problem		
					CT2407	Active implantable	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem		
							F1901	Additional Surgery	A24	No Apparent Device Problem		
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem		
							F1901	Additional Surgery	A24	No Apparent Device Problem		
					CT2727	Active implant devices/ancillaries	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem		
							F1901	Additional Surgery	A24	No Apparent Device Problem		
					CT2750	Implantable cardiac electrical stimulators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem		
							F1901	Additional Surgery	A24	No Apparent Device Problem		
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem		
							F1901	Additional Surgery	A24	No Apparent Device Problem		
					CT2985	Active	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem		
							F1901	Additional Surgery	A24	No Apparent Device Problem		
					CT3001	Implantable metal	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem		
							F1901	Additional Surgery	A24	No Apparent Device Problem		
					CT3002	Implantable synthetic polymer	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem		
							F1901	Additional Surgery	A24	No Apparent Device Problem		
					CT3006	Implantable inorganic material	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem		
							F1901	Additional Surgery	A24	No Apparent Device Problem		
					CT314	Electrically powered	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem		
							F1901	Additional Surgery	A24	No Apparent Device Problem		
					CT321	Long-term surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem		
							F1901	Additional Surgery	A24	No Apparent Device Problem		
					CT449	Generators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem		
							F1901	Additional Surgery	A24	No Apparent Device Problem		
					CT981	Single use	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem		
							F1901	Additional Surgery	A24	No Apparent Device Problem		
					CT983	Surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem		
							F1901	Additional Surgery	A24	No Apparent Device Problem		
13655	2023		47265	Dual-chamber implantable pacemaker, rate-responsive	CT141	Electrodes/Leads and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem		
							F1901	Additional Surgery	A24	No Apparent Device Problem		
					CT145	Cardiovascular devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem		
							F1901	Additional Surgery	A24	No Apparent Device Problem		
					CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem		
							F1901	Additional Surgery	A24	No Apparent Device Problem		
					CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem		
							F1901	Additional Surgery	A24	No Apparent Device Problem		
					CT233	Surgical	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem		
							F1901	Additional Surgery	A24	No Apparent Device Problem		
					CT2406	Implantable	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem		
							F1901	Additional Surgery	A24	No Apparent Device Problem		
					CT2407	Active implantable	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem		
							F1901	Additional Surgery	A24	No Apparent Device Problem		

					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem		
							F1901	Additional Surgery	A24	No Apparent Device Problem		
					CT2727	Active implant devices/ancillaries	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem		
							F1901	Additional Surgery	A24	No Apparent Device Problem		
					CT2750	Implantable cardiac electrical stimulators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem		
							F1901	Additional Surgery	A24	No Apparent Device Problem		
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem		
							F1901	Additional Surgery	A24	No Apparent Device Problem		
					CT2985	Active	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem		
							F1901	Additional Surgery	A24	No Apparent Device Problem		
					CT3001	Implantable metal	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem		
							F1901	Additional Surgery	A24	No Apparent Device Problem		
					CT3002	Implantable synthetic polymer	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem		
							F1901	Additional Surgery	A24	No Apparent Device Problem		
					CT3006	Implantable inorganic material	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem		
							F1901	Additional Surgery	A24	No Apparent Device Problem		
					CT314	Electrically powered	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem		
							F1901	Additional Surgery	A24	No Apparent Device Problem		
					CT321	Long-term surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem		
							F1901	Additional Surgery	A24	No Apparent Device Problem		
					CT449	Generators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem		
							F1901	Additional Surgery	A24	No Apparent Device Problem		
					CT981	Single use	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem		
							F1901	Additional Surgery	A24	No Apparent Device Problem		
					CT983	Surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem		
							F1901	Additional Surgery	A24	No Apparent Device Problem		
13656	2023		47265	Dual-chamber implantable pacemaker, rate-responsive	CT141	Electrodes/Leads and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem		
							F1901	Additional Surgery	A24	No Apparent Device Problem		
					CT145	Cardiovascular devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem		
							F1901	Additional Surgery	A24	No Apparent Device Problem		
					CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem		
							F1901	Additional Surgery	A24	No Apparent Device Problem		
					CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem		
							F1901	Additional Surgery	A24	No Apparent Device Problem		
					CT233	Surgical	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem		
							F1901	Additional Surgery	A24	No Apparent Device Problem		
					CT2406	Implantable	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem		
							F1901	Additional Surgery	A24	No Apparent Device Problem		
					CT2407	Active implantable	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem		
							F1901	Additional Surgery	A24	No Apparent Device Problem		
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem		
							F1901	Additional Surgery	A24	No Apparent Device Problem		
					CT2727	Active implant devices/ancillaries	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem		
							F1901	Additional Surgery	A24	No Apparent Device Problem		
					CT2750	Implantable cardiac electrical stimulators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem		
							F1901	Additional Surgery	A24	No Apparent Device Problem		
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem		
							F1901	Additional Surgery	A24	No Apparent Device Problem		
					CT2985	Active	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem		
							F1901	Additional Surgery	A24	No Apparent Device Problem		
					CT3001	Implantable metal	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem		
							F1901	Additional Surgery	A24	No Apparent Device Problem		
					CT3002	Implantable synthetic polymer	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem		
							F1901	Additional Surgery	A24	No Apparent Device Problem		
					CT3006	Implantable inorganic material	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem		
							F1901	Additional Surgery	A24	No Apparent Device Problem		
					CT314	Electrically powered	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem		
							F1901	Additional Surgery	A24	No Apparent Device Problem		
					CT321	Long-term surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem		
							F1901	Additional Surgery	A24	No Apparent Device Problem		
					CT449	Generators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem		
							F1901	Additional Surgery	A24	No Apparent Device Problem		
					CT981	Single use	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem		
							F1901	Additional Surgery	A24	No Apparent Device Problem		
					CT983	Surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem		
							F1901	Additional Surgery	A24	No Apparent Device Problem		
13657	2023	All other reportable incidents	47263	Cardiac resynchronization therapy implantable pacemaker	CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A0908	Incorrect, Inadequate or Imprecise Result or Readings	12/04/2023	12/04/2023
							F08	Hospitalization or Prolonged Hospitalization	A0908	Incorrect, Inadequate or Imprecise Result or Readings	12/04/2023	12/04/2023
							F1905	Device Revision or Replacement	A0908	Incorrect, Inadequate or Imprecise Result or Readings	12/04/2023	12/04/2023
					CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A0908	Incorrect, Inadequate or Imprecise Result or Readings	12/04/2023	12/04/2023
							F08	Hospitalization or Prolonged Hospitalization	A0908	Incorrect, Inadequate or Imprecise Result or Readings	12/04/2023	12/04/2023
							F1905	Device Revision or Replacement	A0908	Incorrect, Inadequate or Imprecise Result or Readings	12/04/2023	12/04/2023
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A0908	Incorrect, Inadequate or Imprecise Result or Readings	12/04/2023	12/04/2023
							F08	Hospitalization or Prolonged Hospitalization	A0908	Incorrect, Inadequate or Imprecise Result or Readings	12/04/2023	12/04/2023
							F1905	Device Revision or Replacement	A0908	Incorrect, Inadequate or Imprecise Result or Readings	12/04/2023	12/04/2023
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A0908	Incorrect, Inadequate or Imprecise Result or Readings	12/04/2023	12/04/2023
							F08	Hospitalization or Prolonged Hospitalization	A0908	Incorrect, Inadequate or Imprecise Result or Readings	12/04/2023	12/04/2023
							F1905	Device Revision or Replacement	A0908	Incorrect, Inadequate or Imprecise Result or Readings	12/04/2023	12/04/2023
13658	2023	All other reportable incidents	17882	Physiologic-monitoring defibrillation system	CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A07	Electrical /Electronic Property Problem		

									A070803	Failure to Power Up			
							F27	Problem Identified During Non-Clinical Procedure	A07	Electrical /Electronic Property Problem			
									A070803	Failure to Power Up			
13659	2023	All other reportable incidents	17882	Physiologic-monitoring defibrillation system	CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A07	Electrical /Electronic Property Problem			
							F27	Problem Identified During Non-Clinical Procedure	A07	Electrical /Electronic Property Problem			
13660	2023	All other reportable incidents	47263	Cardiac resynchronization therapy implantable pacemaker	CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A0908	Incorrect, Inadequate or Imprecise Result or Readings			
							F26	No Health Consequences or Impact	A0908	Incorrect, Inadequate or Imprecise Result or Readings			
					CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A0908	Incorrect, Inadequate or Imprecise Result or Readings			
							F26	No Health Consequences or Impact	A0908	Incorrect, Inadequate or Imprecise Result or Readings			
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A0908	Incorrect, Inadequate or Imprecise Result or Readings			
							F26	No Health Consequences or Impact	A0908	Incorrect, Inadequate or Imprecise Result or Readings			
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A0908	Incorrect, Inadequate or Imprecise Result or Readings			
							F26	No Health Consequences or Impact	A0908	Incorrect, Inadequate or Imprecise Result or Readings			
13661	2023	All other reportable incidents	47270	Cardiac resynchronization therapy implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A090811	Unable to Obtain Readings			
							F26	No Health Consequences or Impact	A090811	Unable to Obtain Readings			
13662	2023		35223	Endocardial/interventricular septal pacing lead	CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070101	Failure to Capture			
							F26	No Health Consequences or Impact	A070101	Failure to Capture			
13663	2023		47265	Dual-chamber implantable pacemaker, rate-responsive	CT141	Electrodes/Leads and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A0203	Defective Device			
									A0705	Battery Problem			
							F26	No Health Consequences or Impact	A0203	Defective Device			
									A0705	Battery Problem			
					CT145	Cardiovascular devices	E2403	No Clinical Signs, Symptoms or Conditions	A0203	Defective Device			
									A0705	Battery Problem			
							F26	No Health Consequences or Impact	A0203	Defective Device			
									A0705	Battery Problem			
					CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A0203	Defective Device			
									A0705	Battery Problem			
							F26	No Health Consequences or Impact	A0203	Defective Device			
									A0705	Battery Problem			
					CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A0203	Defective Device			
									A0705	Battery Problem			
							F26	No Health Consequences or Impact	A0203	Defective Device			
									A0705	Battery Problem			
					CT233	Surgical	E2403	No Clinical Signs, Symptoms or Conditions	A0203	Defective Device			
									A0705	Battery Problem			
							F26	No Health Consequences or Impact	A0203	Defective Device			
									A0705	Battery Problem			
					CT2406	Implantable	E2403	No Clinical Signs, Symptoms or Conditions	A0203	Defective Device			
									A0705	Battery Problem			
							F26	No Health Consequences or Impact	A0203	Defective Device			
									A0705	Battery Problem			
					CT2407	Active implantable	E2403	No Clinical Signs, Symptoms or Conditions	A0203	Defective Device			
									A0705	Battery Problem			
							F26	No Health Consequences or Impact	A0203	Defective Device			
									A0705	Battery Problem			
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A0203	Defective Device			
									A0705	Battery Problem			
							F26	No Health Consequences or Impact	A0203	Defective Device			
									A0705	Battery Problem			
					CT2727	Active implant devices/ancillaries	E2403	No Clinical Signs, Symptoms or Conditions	A0203	Defective Device			
									A0705	Battery Problem			
							F26	No Health Consequences or Impact	A0203	Defective Device			
									A0705	Battery Problem			
					CT2750	Implantable cardiac electrical stimulators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A0203	Defective Device			
									A0705	Battery Problem			
							F26	No Health Consequences or Impact	A0203	Defective Device			
									A0705	Battery Problem			
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A0203	Defective Device			
									A0705	Battery Problem			
							F26	No Health Consequences or Impact	A0203	Defective Device			
									A0705	Battery Problem			
					CT2985	Active	E2403	No Clinical Signs, Symptoms or Conditions	A0203	Defective Device			
									A0705	Battery Problem			
							F26	No Health Consequences or Impact	A0203	Defective Device			
									A0705	Battery Problem			
					CT3001	Implantable metal	E2403	No Clinical Signs, Symptoms or Conditions	A0203	Defective Device			
									A0705	Battery Problem			
							F26	No Health Consequences or Impact	A0203	Defective Device			
									A0705	Battery Problem			
					CT3002	Implantable synthetic polymer	E2403	No Clinical Signs, Symptoms or Conditions	A0203	Defective Device			
									A0705	Battery Problem			
							F26	No Health Consequences or Impact	A0203	Defective Device			
									A0705	Battery Problem			
					CT3006	Implantable inorganic material	E2403	No Clinical Signs, Symptoms or Conditions	A0203	Defective Device			
									A0705	Battery Problem			
							F26	No Health Consequences or Impact	A0203	Defective Device			
									A0705	Battery Problem			
					CT314	Electrically powered	E2403	No Clinical Signs, Symptoms or Conditions	A0203	Defective Device			
									A0705	Battery Problem			
							F26	No Health Consequences or Impact	A0203	Defective Device			
									A0705	Battery Problem			
					CT321	Long-term surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A0203	Defective Device			

									A0705	Battery Problem		
							F26	No Health Consequences or Impact	A0203	Defective Device		
									A0705	Battery Problem		
					CT449	Generators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A0203	Defective Device		
									A0705	Battery Problem		
							F26	No Health Consequences or Impact	A0203	Defective Device		
									A0705	Battery Problem		
					CT981	Single use	E2403	No Clinical Signs, Symptoms or Conditions	A0203	Defective Device		
									A0705	Battery Problem		
							F26	No Health Consequences or Impact	A0203	Defective Device		
									A0705	Battery Problem		
					CT983	Surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A0203	Defective Device		
									A0705	Battery Problem		
							F26	No Health Consequences or Impact	A0203	Defective Device		
									A0705	Battery Problem		
13664	2023	All other reportable incidents	47265	Dual-chamber implantable pacemaker, rate-responsive	CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A0711	Interrogation Problem	14/03/2025	14/03/2025
									A0908	Incorrect, Inadequate or Imprecise Result or Readings	14/03/2025	14/03/2025
							F08	Hospitalization or Prolonged Hospitalization	A0711	Interrogation Problem	14/03/2025	14/03/2025
									A0908	Incorrect, Inadequate or Imprecise Result or Readings	14/03/2025	14/03/2025
							F1905	Device Revision or Replacement	A0711	Interrogation Problem	14/03/2025	14/03/2025
									A0908	Incorrect, Inadequate or Imprecise Result or Readings	14/03/2025	14/03/2025
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A0711	Interrogation Problem	14/03/2025	14/03/2025
									A0908	Incorrect, Inadequate or Imprecise Result or Readings	14/03/2025	14/03/2025
							F08	Hospitalization or Prolonged Hospitalization	A0711	Interrogation Problem	14/03/2025	14/03/2025
									A0908	Incorrect, Inadequate or Imprecise Result or Readings	14/03/2025	14/03/2025
							F1905	Device Revision or Replacement	A0711	Interrogation Problem	14/03/2025	14/03/2025
									A0908	Incorrect, Inadequate or Imprecise Result or Readings	14/03/2025	14/03/2025
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A0711	Interrogation Problem	14/03/2025	14/03/2025
									A0908	Incorrect, Inadequate or Imprecise Result or Readings	14/03/2025	14/03/2025
							F08	Hospitalization or Prolonged Hospitalization	A0711	Interrogation Problem	14/03/2025	14/03/2025
									A0908	Incorrect, Inadequate or Imprecise Result or Readings	14/03/2025	14/03/2025
							F1905	Device Revision or Replacement	A0711	Interrogation Problem	14/03/2025	14/03/2025
									A0908	Incorrect, Inadequate or Imprecise Result or Readings	14/03/2025	14/03/2025
13665	2023		17882	Physiologic-monitoring defibrillation system	CT1684	Cardiac pacemakers and associated devices	E0602	Cardiac Arrest	A0203	Defective Device		
									A0712	Pacing Problem		
									A071301	Failure to Deliver Shock/Stimulation		
									A0719	Unexpected Shutdown		
									A1104	Computer Operating System Problem		
							F2306	Resuscitation	A0203	Defective Device		
									A0712	Pacing Problem		
									A071301	Failure to Deliver Shock/Stimulation		
									A0719	Unexpected Shutdown		
									A1104	Computer Operating System Problem		
13666	2023	All other reportable incidents	35039	Epicardial pacing lead	CT1684	Cardiac pacemakers and associated devices	E0112	Dizziness	A0712	Pacing Problem		
									A072202	Low Impedance		
									A090801	Signal Artifact/Noise		
							F12	Serious Injury/ Illness/ Impairment	A0712	Pacing Problem		
									A072202	Low Impedance		
									A090801	Signal Artifact/Noise		
							F23	Unexpected Medical Intervention	A0712	Pacing Problem		
									A072202	Low Impedance		
									A090801	Signal Artifact/Noise		
13667	2023	All other reportable incidents	17882	Physiologic-monitoring defibrillation system	CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070803	Failure to Power Up		
							F27	Problem Identified During Non-Clinical Procedure	A070803	Failure to Power Up		
13668	2023		18145	Cardiac myoplasty electrical stimulation system	CT1684	Cardiac pacemakers and associated devices	E24	Other Terms/Codes Related to Clinical Signs and Symptoms or Conditions	A07	Electrical /Electronic Property Problem		
									A0705	Battery Problem		
							F26	No Health Consequences or Impact	A07	Electrical /Electronic Property Problem		
									A0705	Battery Problem		
					CT2866	Implantable cardiac electrical stimulators	E24	Other Terms/Codes Related to Clinical Signs and Symptoms or Conditions	A07	Electrical /Electronic Property Problem		
									A0705	Battery Problem		
							F26	No Health Consequences or Impact	A07	Electrical /Electronic Property Problem		
									A0705	Battery Problem		
13669	2023		47265	Dual-chamber implantable pacemaker, rate-responsive	CT141	Electrodes/Leads and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem		
							F1901	Additional Surgery	A24	No Apparent Device Problem		
					CT145	Cardiovascular devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem		
							F1901	Additional Surgery	A24	No Apparent Device Problem		
					CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem		
							F1901	Additional Surgery	A24	No Apparent Device Problem		
					CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem		
							F1901	Additional Surgery	A24	No Apparent Device Problem		
					CT233	Surgical	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem		
							F1901	Additional Surgery	A24	No Apparent Device Problem		
					CT2406	Implantable	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem		
							F1901	Additional Surgery	A24	No Apparent Device Problem		
					CT2407	Active implantable	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem		
							F1901	Additional Surgery	A24	No Apparent Device Problem		
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem		
							F1901	Additional Surgery	A24	No Apparent Device Problem		
					CT2727	Active implant devices/ancillaries	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem		
							F1901	Additional Surgery	A24	No Apparent Device Problem		
					CT2750	Implantable cardiac electrical stimulators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem		

							F1901	Additional Surgery	A24	No Apparent Device Problem		
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem		
							F1901	Additional Surgery	A24	No Apparent Device Problem		
					CT2985	Active	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem		
							F1901	Additional Surgery	A24	No Apparent Device Problem		
					CT3001	Implantable metal	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem		
							F1901	Additional Surgery	A24	No Apparent Device Problem		
					CT3002	Implantable synthetic polymer	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem		
							F1901	Additional Surgery	A24	No Apparent Device Problem		
					CT3006	Implantable inorganic material	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem		
							F1901	Additional Surgery	A24	No Apparent Device Problem		
					CT314	Electrically powered	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem		
							F1901	Additional Surgery	A24	No Apparent Device Problem		
					CT321	Long-term surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem		
							F1901	Additional Surgery	A24	No Apparent Device Problem		
					CT449	Generators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem		
							F1901	Additional Surgery	A24	No Apparent Device Problem		
					CT981	Single use	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem		
							F1901	Additional Surgery	A24	No Apparent Device Problem		
					CT983	Surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem		
							F1901	Additional Surgery	A24	No Apparent Device Problem		
13670	2023	All other reportable incidents	17882	Physiologic-monitoring defibrillation system	CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A0719	Unexpected Shutdown		
							F27	Problem Identified During Non-Clinical Procedure	A0719	Unexpected Shutdown		
13671	2023		47265	Dual-chamber implantable pacemaker, rate-responsive	CT141	Electrodes/Leads and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	02/02/2023	02/02/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	02/02/2023	02/02/2023
					CT145	Cardiovascular devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	02/02/2023	02/02/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	02/02/2023	02/02/2023
					CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	02/02/2023	02/02/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	02/02/2023	02/02/2023
					CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	02/02/2023	02/02/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	02/02/2023	02/02/2023
					CT233	Surgical	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	02/02/2023	02/02/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	02/02/2023	02/02/2023
					CT2406	Implantable	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	02/02/2023	02/02/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	02/02/2023	02/02/2023
					CT2407	Active implantable	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	02/02/2023	02/02/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	02/02/2023	02/02/2023
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	02/02/2023	02/02/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	02/02/2023	02/02/2023
					CT2727	Active implant devices/ancillaries	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	02/02/2023	02/02/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	02/02/2023	02/02/2023
					CT2750	Implantable cardiac electrical stimulators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	02/02/2023	02/02/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	02/02/2023	02/02/2023
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	02/02/2023	02/02/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	02/02/2023	02/02/2023
					CT2985	Active	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	02/02/2023	02/02/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	02/02/2023	02/02/2023
					CT3001	Implantable metal	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	02/02/2023	02/02/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	02/02/2023	02/02/2023
					CT3002	Implantable synthetic polymer	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	02/02/2023	02/02/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	02/02/2023	02/02/2023
					CT3006	Implantable inorganic material	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	02/02/2023	02/02/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	02/02/2023	02/02/2023
					CT314	Electrically powered	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	02/02/2023	02/02/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	02/02/2023	02/02/2023
					CT321	Long-term surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	02/02/2023	02/02/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	02/02/2023	02/02/2023
					CT449	Generators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	02/02/2023	02/02/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	02/02/2023	02/02/2023
					CT981	Single use	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	02/02/2023	02/02/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	02/02/2023	02/02/2023
					CT983	Surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	02/02/2023	02/02/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	02/02/2023	02/02/2023
13672	2023		47265	Dual-chamber implantable pacemaker, rate-responsive	CT141	Electrodes/Leads and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	08/06/2023	08/06/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	08/06/2023	08/06/2023
					CT145	Cardiovascular devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	08/06/2023	08/06/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	08/06/2023	08/06/2023
					CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	08/06/2023	08/06/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	08/06/2023	08/06/2023
					CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	08/06/2023	08/06/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	08/06/2023	08/06/2023
					CT233	Surgical	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	08/06/2023	08/06/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	08/06/2023	08/06/2023
					CT2406	Implantable	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	08/06/2023	08/06/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	08/06/2023	08/06/2023
					CT2407	Active implantable	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	08/06/2023	08/06/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	08/06/2023	08/06/2023
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	08/06/2023	08/06/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	08/06/2023	08/06/2023
					CT2727	Active implant devices/ancillaries	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	08/06/2023	08/06/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	08/06/2023	08/06/2023
					CT2750	Implantable cardiac electrical stimulators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	08/06/2023	08/06/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	08/06/2023	08/06/2023
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	08/06/2023	08/06/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	08/06/2023	08/06/2023
					CT2985	Active	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	08/06/2023	08/06/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	08/06/2023	08/06/2023
					CT3001	Implantable metal	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	08/06/2023	08/06/2023

						F1901	Additional Surgery	A24	No Apparent Device Problem	08/06/2023	08/06/2023
					CT3002	Implantable synthetic polymer	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	08/06/2023	08/06/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	08/06/2023	08/06/2023
					CT3006	Implantable inorganic material	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	08/06/2023	08/06/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	08/06/2023	08/06/2023
					CT314	Electrically powered	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	08/06/2023	08/06/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	08/06/2023	08/06/2023
					CT321	Long-term surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	08/06/2023	08/06/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	08/06/2023	08/06/2023
					CT449	Generators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	08/06/2023	08/06/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	08/06/2023	08/06/2023
					CT981	Single use	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	08/06/2023	08/06/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	08/06/2023	08/06/2023
					CT983	Surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	08/06/2023	08/06/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	08/06/2023	08/06/2023
13673	2023		47265	Dual-chamber implantable pacemaker, rate-responsive	CT141	Electrodes/Leads and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	09/02/2023	09/02/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	09/02/2023	09/02/2023
					CT145	Cardiovascular devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	09/02/2023	09/02/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	09/02/2023	09/02/2023
					CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	09/02/2023	09/02/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	09/02/2023	09/02/2023
					CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	09/02/2023	09/02/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	09/02/2023	09/02/2023
					CT233	Surgical	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	09/02/2023	09/02/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	09/02/2023	09/02/2023
					CT2406	Implantable	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	09/02/2023	09/02/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	09/02/2023	09/02/2023
					CT2407	Active implantable	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	09/02/2023	09/02/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	09/02/2023	09/02/2023
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	09/02/2023	09/02/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	09/02/2023	09/02/2023
					CT2727	Active implant devices/ancillaries	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	09/02/2023	09/02/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	09/02/2023	09/02/2023
					CT2750	Implantable cardiac electrical stimulators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	09/02/2023	09/02/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	09/02/2023	09/02/2023
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	09/02/2023	09/02/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	09/02/2023	09/02/2023
					CT2985	Active	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	09/02/2023	09/02/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	09/02/2023	09/02/2023
					CT3001	Implantable metal	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	09/02/2023	09/02/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	09/02/2023	09/02/2023
					CT3002	Implantable synthetic polymer	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	09/02/2023	09/02/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	09/02/2023	09/02/2023
					CT3006	Implantable inorganic material	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	09/02/2023	09/02/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	09/02/2023	09/02/2023
					CT314	Electrically powered	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	09/02/2023	09/02/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	09/02/2023	09/02/2023
					CT321	Long-term surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	09/02/2023	09/02/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	09/02/2023	09/02/2023
					CT449	Generators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	09/02/2023	09/02/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	09/02/2023	09/02/2023
					CT981	Single use	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	09/02/2023	09/02/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	09/02/2023	09/02/2023
					CT983	Surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	09/02/2023	09/02/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	09/02/2023	09/02/2023
13674	2023		47265	Dual-chamber implantable pacemaker, rate-responsive	CT141	Electrodes/Leads and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F1901	Additional Surgery	A070504	Premature Discharge of Battery		
					CT145	Cardiovascular devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F1901	Additional Surgery	A070504	Premature Discharge of Battery		
					CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F1901	Additional Surgery	A070504	Premature Discharge of Battery		
					CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F1901	Additional Surgery	A070504	Premature Discharge of Battery		
					CT233	Surgical	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F1901	Additional Surgery	A070504	Premature Discharge of Battery		
					CT2406	Implantable	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F1901	Additional Surgery	A070504	Premature Discharge of Battery		
					CT2407	Active implantable	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F1901	Additional Surgery	A070504	Premature Discharge of Battery		
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F1901	Additional Surgery	A070504	Premature Discharge of Battery		
					CT2727	Active implant devices/ancillaries	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F1901	Additional Surgery	A070504	Premature Discharge of Battery		
					CT2750	Implantable cardiac electrical stimulators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F1901	Additional Surgery	A070504	Premature Discharge of Battery		
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F1901	Additional Surgery	A070504	Premature Discharge of Battery		
					CT2985	Active	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F1901	Additional Surgery	A070504	Premature Discharge of Battery		
					CT3001	Implantable metal	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F1901	Additional Surgery	A070504	Premature Discharge of Battery		
					CT3002	Implantable synthetic polymer	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F1901	Additional Surgery	A070504	Premature Discharge of Battery		
					CT3006	Implantable inorganic material	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F1901	Additional Surgery	A070504	Premature Discharge of Battery		
					CT314	Electrically powered	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F1901	Additional Surgery	A070504	Premature Discharge of Battery		
					CT321	Long-term surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		

							F1901	Additional Surgery	A070504	Premature Discharge of Battery			
					CT449	Generators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery			
							F1901	Additional Surgery	A070504	Premature Discharge of Battery			
					CT981	Single use	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery			
							F1901	Additional Surgery	A070504	Premature Discharge of Battery			
					CT983	Surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery			
							F1901	Additional Surgery	A070504	Premature Discharge of Battery			
13675	2023		47265	Dual-chamber implantable pacemaker, rate-responsive	CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	04/04/2023	04/04/2023	
							F1901	Additional Surgery	A24	No Apparent Device Problem	04/04/2023	04/04/2023	
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	04/04/2023	04/04/2023	
							F1901	Additional Surgery	A24	No Apparent Device Problem	04/04/2023	04/04/2023	
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	04/04/2023	04/04/2023	
							F1901	Additional Surgery	A24	No Apparent Device Problem	04/04/2023	04/04/2023	
13676	2023		47265	Dual-chamber implantable pacemaker, rate-responsive	CT141	Electrodes/Leads and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	02/02/2023	02/02/2023	
							F1901	Additional Surgery	A24	No Apparent Device Problem	02/02/2023	02/02/2023	
					CT145	Cardiovascular devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	02/02/2023	02/02/2023	
							F1901	Additional Surgery	A24	No Apparent Device Problem	02/02/2023	02/02/2023	
					CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	02/02/2023	02/02/2023	
							F1901	Additional Surgery	A24	No Apparent Device Problem	02/02/2023	02/02/2023	
					CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	02/02/2023	02/02/2023	
							F1901	Additional Surgery	A24	No Apparent Device Problem	02/02/2023	02/02/2023	
					CT233	Surgical	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	02/02/2023	02/02/2023	
							F1901	Additional Surgery	A24	No Apparent Device Problem	02/02/2023	02/02/2023	
					CT2406	Implantable	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	02/02/2023	02/02/2023	
							F1901	Additional Surgery	A24	No Apparent Device Problem	02/02/2023	02/02/2023	
					CT2407	Active implantable	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	02/02/2023	02/02/2023	
							F1901	Additional Surgery	A24	No Apparent Device Problem	02/02/2023	02/02/2023	
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	02/02/2023	02/02/2023	
							F1901	Additional Surgery	A24	No Apparent Device Problem	02/02/2023	02/02/2023	
					CT2727	Active implant devices/ancillaries	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	02/02/2023	02/02/2023	
							F1901	Additional Surgery	A24	No Apparent Device Problem	02/02/2023	02/02/2023	
					CT2750	Implantable cardiac electrical stimulators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	02/02/2023	02/02/2023	
							F1901	Additional Surgery	A24	No Apparent Device Problem	02/02/2023	02/02/2023	
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	02/02/2023	02/02/2023	
							F1901	Additional Surgery	A24	No Apparent Device Problem	02/02/2023	02/02/2023	
					CT2985	Active	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	02/02/2023	02/02/2023	
							F1901	Additional Surgery	A24	No Apparent Device Problem	02/02/2023	02/02/2023	
					CT3001	Implantable metal	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	02/02/2023	02/02/2023	
							F1901	Additional Surgery	A24	No Apparent Device Problem	02/02/2023	02/02/2023	
					CT3002	Implantable synthetic polymer	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	02/02/2023	02/02/2023	
							F1901	Additional Surgery	A24	No Apparent Device Problem	02/02/2023	02/02/2023	
					CT3006	Implantable inorganic material	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	02/02/2023	02/02/2023	
							F1901	Additional Surgery	A24	No Apparent Device Problem	02/02/2023	02/02/2023	
					CT314	Electrically powered	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	02/02/2023	02/02/2023	
							F1901	Additional Surgery	A24	No Apparent Device Problem	02/02/2023	02/02/2023	
					CT321	Long-term surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	02/02/2023	02/02/2023	
							F1901	Additional Surgery	A24	No Apparent Device Problem	02/02/2023	02/02/2023	
					CT449	Generators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	02/02/2023	02/02/2023	
							F1901	Additional Surgery	A24	No Apparent Device Problem	02/02/2023	02/02/2023	
					CT981	Single use	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	02/02/2023	02/02/2023	
							F1901	Additional Surgery	A24	No Apparent Device Problem	02/02/2023	02/02/2023	
					CT983	Surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	02/02/2023	02/02/2023	
							F1901	Additional Surgery	A24	No Apparent Device Problem	02/02/2023	02/02/2023	
13677	2023		47265	Dual-chamber implantable pacemaker, rate-responsive	CT141	Electrodes/Leads and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	22/06/2023	22/06/2023	
							F1901	Additional Surgery	A24	No Apparent Device Problem	22/06/2023	22/06/2023	
					CT145	Cardiovascular devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	22/06/2023	22/06/2023	
							F1901	Additional Surgery	A24	No Apparent Device Problem	22/06/2023	22/06/2023	
					CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	22/06/2023	22/06/2023	
							F1901	Additional Surgery	A24	No Apparent Device Problem	22/06/2023	22/06/2023	
					CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	22/06/2023	22/06/2023	
							F1901	Additional Surgery	A24	No Apparent Device Problem	22/06/2023	22/06/2023	
					CT233	Surgical	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	22/06/2023	22/06/2023	
							F1901	Additional Surgery	A24	No Apparent Device Problem	22/06/2023	22/06/2023	
					CT2406	Implantable	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	22/06/2023	22/06/2023	
							F1901	Additional Surgery	A24	No Apparent Device Problem	22/06/2023	22/06/2023	
					CT2407	Active implantable	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	22/06/2023	22/06/2023	
							F1901	Additional Surgery	A24	No Apparent Device Problem	22/06/2023	22/06/2023	
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	22/06/2023	22/06/2023	
							F1901	Additional Surgery	A24	No Apparent Device Problem	22/06/2023	22/06/2023	
					CT2727	Active implant devices/ancillaries	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	22/06/2023	22/06/2023	
							F1901	Additional Surgery	A24	No Apparent Device Problem	22/06/2023	22/06/2023	
					CT2750	Implantable cardiac electrical stimulators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	22/06/2023	22/06/2023	
							F1901	Additional Surgery	A24	No Apparent Device Problem	22/06/2023	22/06/2023	
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	22/06/2023	22/06/2023	
							F1901	Additional Surgery	A24	No Apparent Device Problem	22/06/2023	22/06/2023	
					CT2985	Active	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	22/06/2023	22/06/2023	
							F1901	Additional Surgery	A24	No Apparent Device Problem	22/06/2023	22/06/2023	
					CT3001	Implantable metal	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	22/06/2023	22/06/2023	
							F1901	Additional Surgery	A24	No Apparent Device Problem	22/06/2023	22/06/2023	
					CT3002	Implantable synthetic polymer	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	22/06/2023	22/06/2023	
							F1901	Additional Surgery	A24	No Apparent Device Problem	22/06/2023	22/06/2023	
					CT3006	Implantable inorganic material	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	22/06/2023	22/06/2023	
							F1901	Additional Surgery	A24	No Apparent Device Problem	22/06/2023	22/06/2023	
					CT314	Electrically powered	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	22/06/2023	22/06/2023	
							F1901	Additional Surgery	A24	No Apparent Device Problem	22/06/2023	22/06/2023	
					CT321	Long-term surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	22/06/2023	22/06/2023	
							F1901	Additional Surgery	A24	No Apparent Device Problem	22/06/2023	22/06/2023	
					CT449	Generators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	22/06/2023	22/06/2023	

						F1901	Additional Surgery	A24	No Apparent Device Problem	22/06/2023	22/06/2023
				CT981	Single use	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	22/06/2023	22/06/2023
						F1901	Additional Surgery	A24	No Apparent Device Problem	22/06/2023	22/06/2023
				CT983	Surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	22/06/2023	22/06/2023
						F1901	Additional Surgery	A24	No Apparent Device Problem	22/06/2023	22/06/2023
13678	2023		47265	Dual-chamber implantable pacemaker, rate-responsive	CT141	Electrodes/Leads and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	07/03/2023	07/03/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	07/03/2023	07/03/2023
					CT145	Cardiovascular devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	07/03/2023	07/03/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	07/03/2023	07/03/2023
					CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	07/03/2023	07/03/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	07/03/2023	07/03/2023
					CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	07/03/2023	07/03/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	07/03/2023	07/03/2023
					CT233	Surgical	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	07/03/2023	07/03/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	07/03/2023	07/03/2023
					CT2406	Implantable	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	07/03/2023	07/03/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	07/03/2023	07/03/2023
					CT2407	Active implantable	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	07/03/2023	07/03/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	07/03/2023	07/03/2023
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	07/03/2023	07/03/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	07/03/2023	07/03/2023
					CT2727	Active implant devices/ancillaries	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	07/03/2023	07/03/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	07/03/2023	07/03/2023
					CT2750	Implantable cardiac electrical stimulators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	07/03/2023	07/03/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	07/03/2023	07/03/2023
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	07/03/2023	07/03/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	07/03/2023	07/03/2023
					CT2985	Active	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	07/03/2023	07/03/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	07/03/2023	07/03/2023
					CT3001	Implantable metal	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	07/03/2023	07/03/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	07/03/2023	07/03/2023
					CT3002	Implantable synthetic polymer	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	07/03/2023	07/03/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	07/03/2023	07/03/2023
					CT3006	Implantable inorganic material	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	07/03/2023	07/03/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	07/03/2023	07/03/2023
					CT314	Electrically powered	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	07/03/2023	07/03/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	07/03/2023	07/03/2023
					CT321	Long-term surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	07/03/2023	07/03/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	07/03/2023	07/03/2023
					CT449	Generators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	07/03/2023	07/03/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	07/03/2023	07/03/2023
					CT981	Single use	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	07/03/2023	07/03/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	07/03/2023	07/03/2023
					CT983	Surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	07/03/2023	07/03/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	07/03/2023	07/03/2023
13679	2023		47265	Dual-chamber implantable pacemaker, rate-responsive	CT141	Electrodes/Leads and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	14/03/2023	14/03/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	14/03/2023	14/03/2023
					CT145	Cardiovascular devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	14/03/2023	14/03/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	14/03/2023	14/03/2023
					CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	14/03/2023	14/03/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	14/03/2023	14/03/2023
					CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	14/03/2023	14/03/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	14/03/2023	14/03/2023
					CT233	Surgical	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	14/03/2023	14/03/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	14/03/2023	14/03/2023
					CT2406	Implantable	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	14/03/2023	14/03/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	14/03/2023	14/03/2023
					CT2407	Active implantable	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	14/03/2023	14/03/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	14/03/2023	14/03/2023
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	14/03/2023	14/03/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	14/03/2023	14/03/2023
					CT2727	Active implant devices/ancillaries	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	14/03/2023	14/03/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	14/03/2023	14/03/2023
					CT2750	Implantable cardiac electrical stimulators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	14/03/2023	14/03/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	14/03/2023	14/03/2023
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	14/03/2023	14/03/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	14/03/2023	14/03/2023
					CT2985	Active	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	14/03/2023	14/03/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	14/03/2023	14/03/2023
					CT3001	Implantable metal	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	14/03/2023	14/03/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	14/03/2023	14/03/2023
					CT3002	Implantable synthetic polymer	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	14/03/2023	14/03/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	14/03/2023	14/03/2023
					CT3006	Implantable inorganic material	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	14/03/2023	14/03/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	14/03/2023	14/03/2023
					CT314	Electrically powered	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	14/03/2023	14/03/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	14/03/2023	14/03/2023
					CT321	Long-term surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	14/03/2023	14/03/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	14/03/2023	14/03/2023
					CT449	Generators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	14/03/2023	14/03/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	14/03/2023	14/03/2023
					CT981	Single use	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	14/03/2023	14/03/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	14/03/2023	14/03/2023
					CT983	Surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	14/03/2023	14/03/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	14/03/2023	14/03/2023
13680	2023		47265	Dual-chamber implantable pacemaker, rate-responsive	CT141	Electrodes/Leads and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	03/03/2023	03/03/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	03/03/2023	03/03/2023
					CT145	Cardiovascular devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	03/03/2023	03/03/2023

							F1901	Additional Surgery	A24	No Apparent Device Problem	03/03/2023	03/03/2023
					CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	03/03/2023	03/03/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	03/03/2023	03/03/2023
					CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	03/03/2023	03/03/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	03/03/2023	03/03/2023
					CT233	Surgical	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	03/03/2023	03/03/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	03/03/2023	03/03/2023
					CT2406	Implantable	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	03/03/2023	03/03/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	03/03/2023	03/03/2023
					CT2407	Active implantable	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	03/03/2023	03/03/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	03/03/2023	03/03/2023
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	03/03/2023	03/03/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	03/03/2023	03/03/2023
					CT2727	Active implant devices/ancillaries	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	03/03/2023	03/03/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	03/03/2023	03/03/2023
					CT2750	Implantable cardiac electrical stimulators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	03/03/2023	03/03/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	03/03/2023	03/03/2023
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	03/03/2023	03/03/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	03/03/2023	03/03/2023
					CT2985	Active	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	03/03/2023	03/03/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	03/03/2023	03/03/2023
					CT3001	Implantable metal	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	03/03/2023	03/03/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	03/03/2023	03/03/2023
					CT3002	Implantable synthetic polymer	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	03/03/2023	03/03/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	03/03/2023	03/03/2023
					CT3006	Implantable inorganic material	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	03/03/2023	03/03/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	03/03/2023	03/03/2023
					CT314	Electrically powered	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	03/03/2023	03/03/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	03/03/2023	03/03/2023
					CT321	Long-term surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	03/03/2023	03/03/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	03/03/2023	03/03/2023
					CT449	Generators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	03/03/2023	03/03/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	03/03/2023	03/03/2023
					CT981	Single use	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	03/03/2023	03/03/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	03/03/2023	03/03/2023
					CT983	Surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	03/03/2023	03/03/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	03/03/2023	03/03/2023
13681	2023	All other reportable incidents	35223	Endocardial/interventricular septal pacing lead	CT1684	Cardiac pacemakers and associated devices	E060106	Heart Block	A0701	Capturing Problem	06/09/2023	06/09/2023
									A070101	Failure to Capture	06/09/2023	06/09/2023
									A070104	Unstable Capture Threshold	06/09/2023	06/09/2023
							F08	Hospitalization or Prolonged Hospitalization	A0701	Capturing Problem	06/09/2023	06/09/2023
									A070101	Failure to Capture	06/09/2023	06/09/2023
									A070104	Unstable Capture Threshold	06/09/2023	06/09/2023
							F12	Serious Injury/ Illness/ Impairment	A0701	Capturing Problem	06/09/2023	06/09/2023
									A070101	Failure to Capture	06/09/2023	06/09/2023
									A070104	Unstable Capture Threshold	06/09/2023	06/09/2023
							F23	Unexpected Medical Intervention	A0701	Capturing Problem	06/09/2023	06/09/2023
									A070101	Failure to Capture	06/09/2023	06/09/2023
									A070104	Unstable Capture Threshold	06/09/2023	06/09/2023
13682	2023	All other reportable incidents	35223	Endocardial/interventricular septal pacing lead	CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A05	Mechanical Problem		
									A0701	Capturing Problem		
									A072202	Low Impedance		
							F26	No Health Consequences or Impact	A05	Mechanical Problem		
									A0701	Capturing Problem		
									A072202	Low Impedance		
13683	2023	All other reportable incidents	47265	Dual-chamber implantable pacemaker, rate-responsive	CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem		
							F1901	Additional Surgery	A24	No Apparent Device Problem		
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem		
							F1901	Additional Surgery	A24	No Apparent Device Problem		
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem		
							F1901	Additional Surgery	A24	No Apparent Device Problem		
13684	2023	All other reportable incidents	35039	Epicardial pacing lead	CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A090801	Signal Artifact/Noise		
							F26	No Health Consequences or Impact	A090801	Signal Artifact/Noise		
13685	2023	All other reportable incidents	47265	Dual-chamber implantable pacemaker, rate-responsive	CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	26/02/2023	26/02/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	26/02/2023	26/02/2023
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	26/02/2023	26/02/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	26/02/2023	26/02/2023
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	26/02/2023	26/02/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	26/02/2023	26/02/2023
13686	2023		47265	Dual-chamber implantable pacemaker, rate-responsive	CT141	Electrodes/Leads and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	13/04/2023	13/04/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	13/04/2023	13/04/2023
					CT145	Cardiovascular devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	13/04/2023	13/04/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	13/04/2023	13/04/2023
					CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	13/04/2023	13/04/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	13/04/2023	13/04/2023
					CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	13/04/2023	13/04/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	13/04/2023	13/04/2023
					CT233	Surgical	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	13/04/2023	13/04/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	13/04/2023	13/04/2023
					CT2406	Implantable	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	13/04/2023	13/04/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	13/04/2023	13/04/2023
					CT2407	Active implantable	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	13/04/2023	13/04/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	13/04/2023	13/04/2023
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	13/04/2023	13/04/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	13/04/2023	13/04/2023
					CT2727	Active implant devices/ancillaries	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	13/04/2023	13/04/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	13/04/2023	13/04/2023
					CT2750	Implantable cardiac electrical stimulators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	13/04/2023	13/04/2023

							F1901	Additional Surgery	A24	No Apparent Device Problem	13/04/2023	13/04/2023
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	13/04/2023	13/04/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	13/04/2023	13/04/2023
					CT2985	Active	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	13/04/2023	13/04/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	13/04/2023	13/04/2023
					CT3001	Implantable metal	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	13/04/2023	13/04/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	13/04/2023	13/04/2023
					CT3002	Implantable synthetic polymer	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	13/04/2023	13/04/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	13/04/2023	13/04/2023
					CT3006	Implantable inorganic material	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	13/04/2023	13/04/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	13/04/2023	13/04/2023
					CT314	Electrically powered	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	13/04/2023	13/04/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	13/04/2023	13/04/2023
					CT321	Long-term surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	13/04/2023	13/04/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	13/04/2023	13/04/2023
					CT449	Generators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	13/04/2023	13/04/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	13/04/2023	13/04/2023
					CT981	Single use	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	13/04/2023	13/04/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	13/04/2023	13/04/2023
					CT983	Surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	13/04/2023	13/04/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	13/04/2023	13/04/2023
13687	2023		47265	Dual-chamber implantable pacemaker, rate-responsive	CT141	Electrodes/Leads and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	19/03/2023	19/03/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	19/03/2023	19/03/2023
					CT145	Cardiovascular devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	19/03/2023	19/03/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	19/03/2023	19/03/2023
					CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	19/03/2023	19/03/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	19/03/2023	19/03/2023
					CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	19/03/2023	19/03/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	19/03/2023	19/03/2023
					CT233	Surgical	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	19/03/2023	19/03/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	19/03/2023	19/03/2023
					CT2406	Implantable	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	19/03/2023	19/03/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	19/03/2023	19/03/2023
					CT2407	Active implantable	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	19/03/2023	19/03/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	19/03/2023	19/03/2023
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	19/03/2023	19/03/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	19/03/2023	19/03/2023
					CT2727	Active implant devices/ancillaries	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	19/03/2023	19/03/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	19/03/2023	19/03/2023
					CT2750	Implantable cardiac electrical stimulators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	19/03/2023	19/03/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	19/03/2023	19/03/2023
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	19/03/2023	19/03/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	19/03/2023	19/03/2023
					CT2985	Active	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	19/03/2023	19/03/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	19/03/2023	19/03/2023
					CT3001	Implantable metal	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	19/03/2023	19/03/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	19/03/2023	19/03/2023
					CT3002	Implantable synthetic polymer	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	19/03/2023	19/03/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	19/03/2023	19/03/2023
					CT3006	Implantable inorganic material	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	19/03/2023	19/03/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	19/03/2023	19/03/2023
					CT314	Electrically powered	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	19/03/2023	19/03/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	19/03/2023	19/03/2023
					CT321	Long-term surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	19/03/2023	19/03/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	19/03/2023	19/03/2023
					CT449	Generators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	19/03/2023	19/03/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	19/03/2023	19/03/2023
					CT981	Single use	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	19/03/2023	19/03/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	19/03/2023	19/03/2023
					CT983	Surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	19/03/2023	19/03/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	19/03/2023	19/03/2023
13688	2023		47265	Dual-chamber implantable pacemaker, rate-responsive	CT141	Electrodes/Leads and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	22/06/2023	22/06/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	22/06/2023	22/06/2023
					CT145	Cardiovascular devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	22/06/2023	22/06/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	22/06/2023	22/06/2023
					CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	22/06/2023	22/06/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	22/06/2023	22/06/2023
					CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	22/06/2023	22/06/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	22/06/2023	22/06/2023
					CT233	Surgical	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	22/06/2023	22/06/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	22/06/2023	22/06/2023
					CT2406	Implantable	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	22/06/2023	22/06/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	22/06/2023	22/06/2023
					CT2407	Active implantable	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	22/06/2023	22/06/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	22/06/2023	22/06/2023
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	22/06/2023	22/06/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	22/06/2023	22/06/2023
					CT2727	Active implant devices/ancillaries	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	22/06/2023	22/06/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	22/06/2023	22/06/2023
					CT2750	Implantable cardiac electrical stimulators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	22/06/2023	22/06/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	22/06/2023	22/06/2023
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	22/06/2023	22/06/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	22/06/2023	22/06/2023
					CT2985	Active	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	22/06/2023	22/06/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	22/06/2023	22/06/2023
					CT3001	Implantable metal	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	22/06/2023	22/06/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	22/06/2023	22/06/2023
					CT3002	Implantable synthetic polymer	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	22/06/2023	22/06/2023

						F1901	Additional Surgery	A24	No Apparent Device Problem	22/06/2023	22/06/2023
				CT3006	Implantable inorganic material	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	22/06/2023	22/06/2023
						F1901	Additional Surgery	A24	No Apparent Device Problem	22/06/2023	22/06/2023
				CT314	Electrically powered	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	22/06/2023	22/06/2023
						F1901	Additional Surgery	A24	No Apparent Device Problem	22/06/2023	22/06/2023
				CT321	Long-term surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	22/06/2023	22/06/2023
						F1901	Additional Surgery	A24	No Apparent Device Problem	22/06/2023	22/06/2023
				CT449	Generators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	22/06/2023	22/06/2023
						F1901	Additional Surgery	A24	No Apparent Device Problem	22/06/2023	22/06/2023
				CT981	Single use	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	22/06/2023	22/06/2023
						F1901	Additional Surgery	A24	No Apparent Device Problem	22/06/2023	22/06/2023
				CT983	Surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	22/06/2023	22/06/2023
						F1901	Additional Surgery	A24	No Apparent Device Problem	22/06/2023	22/06/2023
13689	2023	All other reportable incidents	17882	Physiologic-monitoring defibrillation system	CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070803	Failure to Power Up		
							F27	Problem Identified During Non-Clinical Procedure	A070803	Failure to Power Up		
13690	2023	All other reportable incidents	17882	Physiologic-monitoring defibrillation system	CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A0719	Unexpected Shutdown		
							F27	Problem Identified During Non-Clinical Procedure	A0719	Unexpected Shutdown		
13691	2023	All other reportable incidents	47265	Dual-chamber implantable pacemaker, rate-responsive	CT141	Electrodes/Leads and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem		
							F1901	Additional Surgery	A24	No Apparent Device Problem		
					CT145	Cardiovascular devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem		
							F1901	Additional Surgery	A24	No Apparent Device Problem		
					CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem		
							F1901	Additional Surgery	A24	No Apparent Device Problem		
					CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem		
							F1901	Additional Surgery	A24	No Apparent Device Problem		
					CT233	Surgical	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem		
							F1901	Additional Surgery	A24	No Apparent Device Problem		
					CT2406	Implantable	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem		
							F1901	Additional Surgery	A24	No Apparent Device Problem		
					CT2407	Active implantable	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem		
							F1901	Additional Surgery	A24	No Apparent Device Problem		
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem		
							F1901	Additional Surgery	A24	No Apparent Device Problem		
					CT2727	Active implant devices/ancillaries	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem		
							F1901	Additional Surgery	A24	No Apparent Device Problem		
					CT2750	Implantable cardiac electrical stimulators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem		
							F1901	Additional Surgery	A24	No Apparent Device Problem		
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem		
							F1901	Additional Surgery	A24	No Apparent Device Problem		
					CT2985	Active	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem		
							F1901	Additional Surgery	A24	No Apparent Device Problem		
					CT3001	Implantable metal	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem		
							F1901	Additional Surgery	A24	No Apparent Device Problem		
					CT3002	Implantable synthetic polymer	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem		
							F1901	Additional Surgery	A24	No Apparent Device Problem		
					CT3006	Implantable inorganic material	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem		
							F1901	Additional Surgery	A24	No Apparent Device Problem		
					CT314	Electrically powered	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem		
							F1901	Additional Surgery	A24	No Apparent Device Problem		
					CT321	Long-term surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem		
							F1901	Additional Surgery	A24	No Apparent Device Problem		
					CT449	Generators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem		
							F1901	Additional Surgery	A24	No Apparent Device Problem		
					CT981	Single use	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem		
							F1901	Additional Surgery	A24	No Apparent Device Problem		
					CT983	Surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem		
							F1901	Additional Surgery	A24	No Apparent Device Problem		
13692	2023		47265	Dual-chamber implantable pacemaker, rate-responsive	CT141	Electrodes/Leads and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	20/06/2023	20/06/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	20/06/2023	20/06/2023
					CT145	Cardiovascular devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	20/06/2023	20/06/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	20/06/2023	20/06/2023
					CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	20/06/2023	20/06/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	20/06/2023	20/06/2023
					CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	20/06/2023	20/06/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	20/06/2023	20/06/2023
					CT233	Surgical	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	20/06/2023	20/06/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	20/06/2023	20/06/2023
					CT2406	Implantable	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	20/06/2023	20/06/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	20/06/2023	20/06/2023
					CT2407	Active implantable	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	20/06/2023	20/06/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	20/06/2023	20/06/2023
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	20/06/2023	20/06/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	20/06/2023	20/06/2023
					CT2727	Active implant devices/ancillaries	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	20/06/2023	20/06/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	20/06/2023	20/06/2023
					CT2750	Implantable cardiac electrical stimulators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	20/06/2023	20/06/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	20/06/2023	20/06/2023
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	20/06/2023	20/06/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	20/06/2023	20/06/2023
					CT2985	Active	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	20/06/2023	20/06/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	20/06/2023	20/06/2023
					CT3001	Implantable metal	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	20/06/2023	20/06/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	20/06/2023	20/06/2023
					CT3002	Implantable synthetic polymer	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	20/06/2023	20/06/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	20/06/2023	20/06/2023
					CT3006	Implantable inorganic material	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	20/06/2023	20/06/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	20/06/2023	20/06/2023
					CT314	Electrically powered	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	20/06/2023	20/06/2023

							F1901	Additional Surgery	A24	No Apparent Device Problem	20/06/2023	20/06/2023
					CT321	Long-term surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	20/06/2023	20/06/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	20/06/2023	20/06/2023
					CT449	Generators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	20/06/2023	20/06/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	20/06/2023	20/06/2023
					CT981	Single use	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	20/06/2023	20/06/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	20/06/2023	20/06/2023
					CT983	Surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	20/06/2023	20/06/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	20/06/2023	20/06/2023
13693	2023		47265	Dual-chamber implantable pacemaker, rate-responsive	CT141	Electrodes/Leads and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	22/06/2023	22/06/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	22/06/2023	22/06/2023
					CT145	Cardiovascular devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	22/06/2023	22/06/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	22/06/2023	22/06/2023
					CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	22/06/2023	22/06/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	22/06/2023	22/06/2023
					CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	22/06/2023	22/06/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	22/06/2023	22/06/2023
					CT233	Surgical	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	22/06/2023	22/06/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	22/06/2023	22/06/2023
					CT2406	Implantable	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	22/06/2023	22/06/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	22/06/2023	22/06/2023
					CT2407	Active implantable	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	22/06/2023	22/06/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	22/06/2023	22/06/2023
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	22/06/2023	22/06/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	22/06/2023	22/06/2023
					CT2727	Active implant devices/ancillaries	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	22/06/2023	22/06/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	22/06/2023	22/06/2023
					CT2750	Implantable cardiac electrical stimulators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	22/06/2023	22/06/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	22/06/2023	22/06/2023
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	22/06/2023	22/06/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	22/06/2023	22/06/2023
					CT2985	Active	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	22/06/2023	22/06/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	22/06/2023	22/06/2023
					CT3001	Implantable metal	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	22/06/2023	22/06/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	22/06/2023	22/06/2023
					CT3002	Implantable synthetic polymer	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	22/06/2023	22/06/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	22/06/2023	22/06/2023
					CT3006	Implantable inorganic material	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	22/06/2023	22/06/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	22/06/2023	22/06/2023
					CT314	Electrically powered	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	22/06/2023	22/06/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	22/06/2023	22/06/2023
					CT321	Long-term surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	22/06/2023	22/06/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	22/06/2023	22/06/2023
					CT449	Generators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	22/06/2023	22/06/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	22/06/2023	22/06/2023
					CT981	Single use	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	22/06/2023	22/06/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	22/06/2023	22/06/2023
					CT983	Surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	22/06/2023	22/06/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	22/06/2023	22/06/2023
13694	2023		47265	Dual-chamber implantable pacemaker, rate-responsive	CT141	Electrodes/Leads and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	22/06/2023	22/06/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	22/06/2023	22/06/2023
					CT145	Cardiovascular devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	22/06/2023	22/06/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	22/06/2023	22/06/2023
					CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	22/06/2023	22/06/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	22/06/2023	22/06/2023
					CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	22/06/2023	22/06/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	22/06/2023	22/06/2023
					CT233	Surgical	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	22/06/2023	22/06/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	22/06/2023	22/06/2023
					CT2406	Implantable	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	22/06/2023	22/06/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	22/06/2023	22/06/2023
					CT2407	Active implantable	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	22/06/2023	22/06/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	22/06/2023	22/06/2023
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	22/06/2023	22/06/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	22/06/2023	22/06/2023
					CT2727	Active implant devices/ancillaries	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	22/06/2023	22/06/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	22/06/2023	22/06/2023
					CT2750	Implantable cardiac electrical stimulators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	22/06/2023	22/06/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	22/06/2023	22/06/2023
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	22/06/2023	22/06/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	22/06/2023	22/06/2023
					CT2985	Active	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	22/06/2023	22/06/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	22/06/2023	22/06/2023
					CT3001	Implantable metal	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	22/06/2023	22/06/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	22/06/2023	22/06/2023
					CT3002	Implantable synthetic polymer	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	22/06/2023	22/06/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	22/06/2023	22/06/2023
					CT3006	Implantable inorganic material	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	22/06/2023	22/06/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	22/06/2023	22/06/2023
					CT314	Electrically powered	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	22/06/2023	22/06/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	22/06/2023	22/06/2023
					CT321	Long-term surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	22/06/2023	22/06/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	22/06/2023	22/06/2023
					CT449	Generators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	22/06/2023	22/06/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	22/06/2023	22/06/2023
					CT981	Single use	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	22/06/2023	22/06/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	22/06/2023	22/06/2023
					CT983	Surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	22/06/2023	22/06/2023

							F1901	Additional Surgery	A24	No Apparent Device Problem	22/06/2023	22/06/2023
13695	2023	All other reportable incidents	47265	Dual-chamber implantable pacemaker, rate-responsive	CT141	Electrodes/Leads and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A090805	Incorrect Measurement	01/01/2023	31/12/2023
							F08	Hospitalization or Prolonged Hospitalization	A090805	Incorrect Measurement	01/01/2023	31/12/2023
							F1905	Device Revision or Replacement	A090805	Incorrect Measurement	01/01/2023	31/12/2023
							F26	No Health Consequences or Impact	A090805	Incorrect Measurement	01/01/2023	31/12/2023
					CT145	Cardiovascular devices	E2403	No Clinical Signs, Symptoms or Conditions	A090805	Incorrect Measurement	01/01/2023	31/12/2023
							F08	Hospitalization or Prolonged Hospitalization	A090805	Incorrect Measurement	01/01/2023	31/12/2023
							F1905	Device Revision or Replacement	A090805	Incorrect Measurement	01/01/2023	31/12/2023
							F26	No Health Consequences or Impact	A090805	Incorrect Measurement	01/01/2023	31/12/2023
					CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A090805	Incorrect Measurement	01/01/2023	31/12/2023
							F08	Hospitalization or Prolonged Hospitalization	A090805	Incorrect Measurement	01/01/2023	31/12/2023
							F1905	Device Revision or Replacement	A090805	Incorrect Measurement	01/01/2023	31/12/2023
							F26	No Health Consequences or Impact	A090805	Incorrect Measurement	01/01/2023	31/12/2023
					CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A090805	Incorrect Measurement	01/01/2023	31/12/2023
							F08	Hospitalization or Prolonged Hospitalization	A090805	Incorrect Measurement	01/01/2023	31/12/2023
							F1905	Device Revision or Replacement	A090805	Incorrect Measurement	01/01/2023	31/12/2023
							F26	No Health Consequences or Impact	A090805	Incorrect Measurement	01/01/2023	31/12/2023
					CT233	Surgical	E2403	No Clinical Signs, Symptoms or Conditions	A090805	Incorrect Measurement	01/01/2023	31/12/2023
							F08	Hospitalization or Prolonged Hospitalization	A090805	Incorrect Measurement	01/01/2023	31/12/2023
							F1905	Device Revision or Replacement	A090805	Incorrect Measurement	01/01/2023	31/12/2023
							F26	No Health Consequences or Impact	A090805	Incorrect Measurement	01/01/2023	31/12/2023
					CT2406	Implantable	E2403	No Clinical Signs, Symptoms or Conditions	A090805	Incorrect Measurement	01/01/2023	31/12/2023
							F08	Hospitalization or Prolonged Hospitalization	A090805	Incorrect Measurement	01/01/2023	31/12/2023
							F1905	Device Revision or Replacement	A090805	Incorrect Measurement	01/01/2023	31/12/2023
							F26	No Health Consequences or Impact	A090805	Incorrect Measurement	01/01/2023	31/12/2023
					CT2407	Active implantable	E2403	No Clinical Signs, Symptoms or Conditions	A090805	Incorrect Measurement	01/01/2023	31/12/2023
							F08	Hospitalization or Prolonged Hospitalization	A090805	Incorrect Measurement	01/01/2023	31/12/2023
							F1905	Device Revision or Replacement	A090805	Incorrect Measurement	01/01/2023	31/12/2023
							F26	No Health Consequences or Impact	A090805	Incorrect Measurement	01/01/2023	31/12/2023
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A090805	Incorrect Measurement	01/01/2023	31/12/2023
							F08	Hospitalization or Prolonged Hospitalization	A090805	Incorrect Measurement	01/01/2023	31/12/2023
							F1905	Device Revision or Replacement	A090805	Incorrect Measurement	01/01/2023	31/12/2023
							F26	No Health Consequences or Impact	A090805	Incorrect Measurement	01/01/2023	31/12/2023
					CT2727	Active implant devices/ancillaries	E2403	No Clinical Signs, Symptoms or Conditions	A090805	Incorrect Measurement	01/01/2023	31/12/2023
							F08	Hospitalization or Prolonged Hospitalization	A090805	Incorrect Measurement	01/01/2023	31/12/2023
							F1905	Device Revision or Replacement	A090805	Incorrect Measurement	01/01/2023	31/12/2023
							F26	No Health Consequences or Impact	A090805	Incorrect Measurement	01/01/2023	31/12/2023
					CT2750	Implantable cardiac electrical stimulators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A090805	Incorrect Measurement	01/01/2023	31/12/2023
							F08	Hospitalization or Prolonged Hospitalization	A090805	Incorrect Measurement	01/01/2023	31/12/2023
							F1905	Device Revision or Replacement	A090805	Incorrect Measurement	01/01/2023	31/12/2023
							F26	No Health Consequences or Impact	A090805	Incorrect Measurement	01/01/2023	31/12/2023
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A090805	Incorrect Measurement	01/01/2023	31/12/2023
							F08	Hospitalization or Prolonged Hospitalization	A090805	Incorrect Measurement	01/01/2023	31/12/2023
							F1905	Device Revision or Replacement	A090805	Incorrect Measurement	01/01/2023	31/12/2023
							F26	No Health Consequences or Impact	A090805	Incorrect Measurement	01/01/2023	31/12/2023
					CT2985	Active	E2403	No Clinical Signs, Symptoms or Conditions	A090805	Incorrect Measurement	01/01/2023	31/12/2023
							F08	Hospitalization or Prolonged Hospitalization	A090805	Incorrect Measurement	01/01/2023	31/12/2023
							F1905	Device Revision or Replacement	A090805	Incorrect Measurement	01/01/2023	31/12/2023
							F26	No Health Consequences or Impact	A090805	Incorrect Measurement	01/01/2023	31/12/2023
					CT3001	Implantable metal	E2403	No Clinical Signs, Symptoms or Conditions	A090805	Incorrect Measurement	01/01/2023	31/12/2023
							F08	Hospitalization or Prolonged Hospitalization	A090805	Incorrect Measurement	01/01/2023	31/12/2023
							F1905	Device Revision or Replacement	A090805	Incorrect Measurement	01/01/2023	31/12/2023
							F26	No Health Consequences or Impact	A090805	Incorrect Measurement	01/01/2023	31/12/2023
					CT3002	Implantable synthetic polymer	E2403	No Clinical Signs, Symptoms or Conditions	A090805	Incorrect Measurement	01/01/2023	31/12/2023
							F08	Hospitalization or Prolonged Hospitalization	A090805	Incorrect Measurement	01/01/2023	31/12/2023
							F1905	Device Revision or Replacement	A090805	Incorrect Measurement	01/01/2023	31/12/2023
							F26	No Health Consequences or Impact	A090805	Incorrect Measurement	01/01/2023	31/12/2023
					CT3006	Implantable inorganic material	E2403	No Clinical Signs, Symptoms or Conditions	A090805	Incorrect Measurement	01/01/2023	31/12/2023
							F08	Hospitalization or Prolonged Hospitalization	A090805	Incorrect Measurement	01/01/2023	31/12/2023
							F1905	Device Revision or Replacement	A090805	Incorrect Measurement	01/01/2023	31/12/2023
							F26	No Health Consequences or Impact	A090805	Incorrect Measurement	01/01/2023	31/12/2023
					CT314	Electrically powered	E2403	No Clinical Signs, Symptoms or Conditions	A090805	Incorrect Measurement	01/01/2023	31/12/2023
							F08	Hospitalization or Prolonged Hospitalization	A090805	Incorrect Measurement	01/01/2023	31/12/2023
							F1905	Device Revision or Replacement	A090805	Incorrect Measurement	01/01/2023	31/12/2023
							F26	No Health Consequences or Impact	A090805	Incorrect Measurement	01/01/2023	31/12/2023
					CT321	Long-term surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A090805	Incorrect Measurement	01/01/2023	31/12/2023
							F08	Hospitalization or Prolonged Hospitalization	A090805	Incorrect Measurement	01/01/2023	31/12/2023
							F1905	Device Revision or Replacement	A090805	Incorrect Measurement	01/01/2023	31/12/2023
							F26	No Health Consequences or Impact	A090805	Incorrect Measurement	01/01/2023	31/12/2023
					CT449	Generators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A090805	Incorrect Measurement	01/01/2023	31/12/2023
							F08	Hospitalization or Prolonged Hospitalization	A090805	Incorrect Measurement	01/01/2023	31/12/2023
							F1905	Device Revision or Replacement	A090805	Incorrect Measurement	01/01/2023	31/12/2023
							F26	No Health Consequences or Impact	A090805	Incorrect Measurement	01/01/2023	31/12/2023
					CT981	Single use	E2403	No Clinical Signs, Symptoms or Conditions	A090805	Incorrect Measurement	01/01/2023	31/12/2023
							F08	Hospitalization or Prolonged Hospitalization	A090805	Incorrect Measurement	01/01/2023	31/12/2023
							F1905	Device Revision or Replacement	A090805	Incorrect Measurement	01/01/2023	31/12/2023
							F26	No Health Consequences or Impact	A090805	Incorrect Measurement	01/01/2023	31/12/2023
					CT983	Surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A090805	Incorrect Measurement	01/01/2023	31/12/2023
							F08	Hospitalization or Prolonged Hospitalization	A090805	Incorrect Measurement	01/01/2023	31/12/2023
							F1905	Device Revision or Replacement	A090805	Incorrect Measurement	01/01/2023	31/12/2023
							F26	No Health Consequences or Impact	A090805	Incorrect Measurement	01/01/2023	31/12/2023
13696	2022	All other reportable incidents	47263	Cardiac resynchronization therapy implantable pacemaker	CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F1901	Additional Surgery	A070504	Premature Discharge of Battery		
					CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F1901	Additional Surgery	A070504	Premature Discharge of Battery		
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F1901	Additional Surgery	A070504	Premature Discharge of Battery		
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		

							F1901	Additional Surgery	A070504	Premature Discharge of Battery		
13697	2022	All other reportable incidents	37265	Dual-chamber implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F1901	Additional Surgery	A070504	Premature Discharge of Battery		
13698	2022	All other reportable incidents	35223	Endocardial/interventricular septal pacing lead	CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070909	Oversensing		
							F1901	Additional Surgery	A070909	Oversensing		
13699	2022	All other reportable incidents	47265	Dual-chamber implantable pacemaker, rate-responsive	CT141	Electrodes/Leads and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	15/08/2022	15/08/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	15/08/2022	15/08/2022
					CT145	Cardiovascular devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	15/08/2022	15/08/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	15/08/2022	15/08/2022
					CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	15/08/2022	15/08/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	15/08/2022	15/08/2022
					CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	15/08/2022	15/08/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	15/08/2022	15/08/2022
					CT233	Surgical	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	15/08/2022	15/08/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	15/08/2022	15/08/2022
					CT2406	Implantable	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	15/08/2022	15/08/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	15/08/2022	15/08/2022
					CT2407	Active implantable	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	15/08/2022	15/08/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	15/08/2022	15/08/2022
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	15/08/2022	15/08/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	15/08/2022	15/08/2022
					CT2727	Active implant devices/ancillaries	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	15/08/2022	15/08/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	15/08/2022	15/08/2022
					CT2750	Implantable cardiac electrical stimulators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	15/08/2022	15/08/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	15/08/2022	15/08/2022
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	15/08/2022	15/08/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	15/08/2022	15/08/2022
					CT2985	Active	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	15/08/2022	15/08/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	15/08/2022	15/08/2022
					CT3001	Implantable metal	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	15/08/2022	15/08/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	15/08/2022	15/08/2022
					CT3002	Implantable synthetic polymer	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	15/08/2022	15/08/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	15/08/2022	15/08/2022
					CT3006	Implantable inorganic material	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	15/08/2022	15/08/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	15/08/2022	15/08/2022
					CT314	Electrically powered	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	15/08/2022	15/08/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	15/08/2022	15/08/2022
					CT321	Long-term surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	15/08/2022	15/08/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	15/08/2022	15/08/2022
					CT449	Generators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	15/08/2022	15/08/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	15/08/2022	15/08/2022
					CT981	Single use	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	15/08/2022	15/08/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	15/08/2022	15/08/2022
					CT983	Surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	15/08/2022	15/08/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	15/08/2022	15/08/2022
13700	2022	All other reportable incidents	35223	Endocardial/interventricular septal pacing lead	CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070909	Oversensing		
							F1908	Prolonged Surgery	A070909	Oversensing		
13701	2022	All other reportable incidents	47205	Cardiac pulse generator programmer	CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A090805	Incorrect Measurement		
							F23	Unexpected Medical Intervention	A090805	Incorrect Measurement		
13702	2022	All other reportable incidents	47265	Dual-chamber implantable pacemaker, rate-responsive	CT1685	Cardiac pacemakers	E011903	Syncope/Fainting	A0712	Pacing Problem	13/04/2022	13/04/2022
							E0601	Arrhythmia	A0712	Pacing Problem	13/04/2022	13/04/2022
							F1908	Prolonged Surgery	A0712	Pacing Problem	13/04/2022	13/04/2022
					CT257	Implantable cardiac pacemakers	E011903	Syncope/Fainting	A0712	Pacing Problem	13/04/2022	13/04/2022
							E0601	Arrhythmia	A0712	Pacing Problem	13/04/2022	13/04/2022
							F1908	Prolonged Surgery	A0712	Pacing Problem	13/04/2022	13/04/2022
					CT2866	Implantable cardiac electrical stimulators	E011903	Syncope/Fainting	A0712	Pacing Problem	13/04/2022	13/04/2022
							E0601	Arrhythmia	A0712	Pacing Problem	13/04/2022	13/04/2022
							F1908	Prolonged Surgery	A0712	Pacing Problem	13/04/2022	13/04/2022
13703	2022	All other reportable incidents	47265	Dual-chamber implantable pacemaker, rate-responsive	CT141	Electrodes/Leads and associated devices	E0601	Arrhythmia	A070101	Failure to Capture	04/04/2022	04/04/2022
									A0712	Pacing Problem	04/04/2022	04/04/2022
									A2303	Improper or Incorrect Procedure or Method	04/04/2022	04/04/2022
							F23	Unexpected Medical Intervention	A070101	Failure to Capture	04/04/2022	04/04/2022
									A0712	Pacing Problem	04/04/2022	04/04/2022
									A2303	Improper or Incorrect Procedure or Method	04/04/2022	04/04/2022
					CT145	Cardiovascular devices	E0601	Arrhythmia	A070101	Failure to Capture	04/04/2022	04/04/2022
									A0712	Pacing Problem	04/04/2022	04/04/2022
									A2303	Improper or Incorrect Procedure or Method	04/04/2022	04/04/2022
							F23	Unexpected Medical Intervention	A070101	Failure to Capture	04/04/2022	04/04/2022
									A0712	Pacing Problem	04/04/2022	04/04/2022
									A2303	Improper or Incorrect Procedure or Method	04/04/2022	04/04/2022
					CT1684	Cardiac pacemakers and associated devices	E0601	Arrhythmia	A070101	Failure to Capture	04/04/2022	04/04/2022
									A0712	Pacing Problem	04/04/2022	04/04/2022
									A2303	Improper or Incorrect Procedure or Method	04/04/2022	04/04/2022
							F23	Unexpected Medical Intervention	A070101	Failure to Capture	04/04/2022	04/04/2022
									A0712	Pacing Problem	04/04/2022	04/04/2022
									A2303	Improper or Incorrect Procedure or Method	04/04/2022	04/04/2022
					CT1685	Cardiac pacemakers	E0601	Arrhythmia	A070101	Failure to Capture	04/04/2022	04/04/2022
									A0712	Pacing Problem	04/04/2022	04/04/2022
									A2303	Improper or Incorrect Procedure or Method	04/04/2022	04/04/2022
							F23	Unexpected Medical Intervention	A070101	Failure to Capture	04/04/2022	04/04/2022
									A0712	Pacing Problem	04/04/2022	04/04/2022
									A2303	Improper or Incorrect Procedure or Method	04/04/2022	04/04/2022
					CT233	Surgical	E0601	Arrhythmia	A070101	Failure to Capture	04/04/2022	04/04/2022
									A0712	Pacing Problem	04/04/2022	04/04/2022
									A2303	Improper or Incorrect Procedure or Method	04/04/2022	04/04/2022
							F23	Unexpected Medical Intervention	A070101	Failure to Capture	04/04/2022	04/04/2022
									A0712	Pacing Problem	04/04/2022	04/04/2022
									A2303	Improper or Incorrect Procedure or Method	04/04/2022	04/04/2022

CT2406	Implantable	E0601	Arrhythmia	A070101	Failure to Capture	04/04/2022	04/04/2022
				A0712	Pacing Problem	04/04/2022	04/04/2022
				A2303	Improper or Incorrect Procedure or Method	04/04/2022	04/04/2022
		F23	Unexpected Medical Intervention	A070101	Failure to Capture	04/04/2022	04/04/2022
				A0712	Pacing Problem	04/04/2022	04/04/2022
				A2303	Improper or Incorrect Procedure or Method	04/04/2022	04/04/2022
CT2407	Active implantable	E0601	Arrhythmia	A070101	Failure to Capture	04/04/2022	04/04/2022
				A0712	Pacing Problem	04/04/2022	04/04/2022
				A2303	Improper or Incorrect Procedure or Method	04/04/2022	04/04/2022
		F23	Unexpected Medical Intervention	A070101	Failure to Capture	04/04/2022	04/04/2022
				A0712	Pacing Problem	04/04/2022	04/04/2022
				A2303	Improper or Incorrect Procedure or Method	04/04/2022	04/04/2022
CT257	Implantable cardiac pacemakers	E0601	Arrhythmia	A070101	Failure to Capture	04/04/2022	04/04/2022
				A0712	Pacing Problem	04/04/2022	04/04/2022
				A2303	Improper or Incorrect Procedure or Method	04/04/2022	04/04/2022
		F23	Unexpected Medical Intervention	A070101	Failure to Capture	04/04/2022	04/04/2022
				A0712	Pacing Problem	04/04/2022	04/04/2022
				A2303	Improper or Incorrect Procedure or Method	04/04/2022	04/04/2022
CT2727	Active implant devices/ancillaries	E0601	Arrhythmia	A070101	Failure to Capture	04/04/2022	04/04/2022
				A0712	Pacing Problem	04/04/2022	04/04/2022
				A2303	Improper or Incorrect Procedure or Method	04/04/2022	04/04/2022
		F23	Unexpected Medical Intervention	A070101	Failure to Capture	04/04/2022	04/04/2022
				A0712	Pacing Problem	04/04/2022	04/04/2022
				A2303	Improper or Incorrect Procedure or Method	04/04/2022	04/04/2022
CT2750	Implantable cardiac electrical stimulators and associated devices	E0601	Arrhythmia	A070101	Failure to Capture	04/04/2022	04/04/2022
				A0712	Pacing Problem	04/04/2022	04/04/2022
				A2303	Improper or Incorrect Procedure or Method	04/04/2022	04/04/2022
		F23	Unexpected Medical Intervention	A070101	Failure to Capture	04/04/2022	04/04/2022
				A0712	Pacing Problem	04/04/2022	04/04/2022
				A2303	Improper or Incorrect Procedure or Method	04/04/2022	04/04/2022
CT2866	Implantable cardiac electrical stimulators	E0601	Arrhythmia	A070101	Failure to Capture	04/04/2022	04/04/2022
				A0712	Pacing Problem	04/04/2022	04/04/2022
				A2303	Improper or Incorrect Procedure or Method	04/04/2022	04/04/2022
		F23	Unexpected Medical Intervention	A070101	Failure to Capture	04/04/2022	04/04/2022
				A0712	Pacing Problem	04/04/2022	04/04/2022
				A2303	Improper or Incorrect Procedure or Method	04/04/2022	04/04/2022
CT2985	Active	E0601	Arrhythmia	A070101	Failure to Capture	04/04/2022	04/04/2022
				A0712	Pacing Problem	04/04/2022	04/04/2022
				A2303	Improper or Incorrect Procedure or Method	04/04/2022	04/04/2022
		F23	Unexpected Medical Intervention	A070101	Failure to Capture	04/04/2022	04/04/2022
				A0712	Pacing Problem	04/04/2022	04/04/2022
				A2303	Improper or Incorrect Procedure or Method	04/04/2022	04/04/2022
CT3001	Implantable metal	E0601	Arrhythmia	A070101	Failure to Capture	04/04/2022	04/04/2022
				A0712	Pacing Problem	04/04/2022	04/04/2022
				A2303	Improper or Incorrect Procedure or Method	04/04/2022	04/04/2022
		F23	Unexpected Medical Intervention	A070101	Failure to Capture	04/04/2022	04/04/2022
				A0712	Pacing Problem	04/04/2022	04/04/2022
				A2303	Improper or Incorrect Procedure or Method	04/04/2022	04/04/2022
CT3002	Implantable synthetic polymer	E0601	Arrhythmia	A070101	Failure to Capture	04/04/2022	04/04/2022
				A0712	Pacing Problem	04/04/2022	04/04/2022
				A2303	Improper or Incorrect Procedure or Method	04/04/2022	04/04/2022
		F23	Unexpected Medical Intervention	A070101	Failure to Capture	04/04/2022	04/04/2022
				A0712	Pacing Problem	04/04/2022	04/04/2022
				A2303	Improper or Incorrect Procedure or Method	04/04/2022	04/04/2022
CT3006	Implantable inorganic material	E0601	Arrhythmia	A070101	Failure to Capture	04/04/2022	04/04/2022
				A0712	Pacing Problem	04/04/2022	04/04/2022
				A2303	Improper or Incorrect Procedure or Method	04/04/2022	04/04/2022
		F23	Unexpected Medical Intervention	A070101	Failure to Capture	04/04/2022	04/04/2022
				A0712	Pacing Problem	04/04/2022	04/04/2022
				A2303	Improper or Incorrect Procedure or Method	04/04/2022	04/04/2022
CT314	Electrically powered	E0601	Arrhythmia	A070101	Failure to Capture	04/04/2022	04/04/2022
				A0712	Pacing Problem	04/04/2022	04/04/2022
				A2303	Improper or Incorrect Procedure or Method	04/04/2022	04/04/2022
		F23	Unexpected Medical Intervention	A070101	Failure to Capture	04/04/2022	04/04/2022
				A0712	Pacing Problem	04/04/2022	04/04/2022
				A2303	Improper or Incorrect Procedure or Method	04/04/2022	04/04/2022
CT321	Long-term surgical invasive	E0601	Arrhythmia	A070101	Failure to Capture	04/04/2022	04/04/2022
				A0712	Pacing Problem	04/04/2022	04/04/2022
				A2303	Improper or Incorrect Procedure or Method	04/04/2022	04/04/2022
		F23	Unexpected Medical Intervention	A070101	Failure to Capture	04/04/2022	04/04/2022
				A0712	Pacing Problem	04/04/2022	04/04/2022
				A2303	Improper or Incorrect Procedure or Method	04/04/2022	04/04/2022
CT449	Generators and associated devices	E0601	Arrhythmia	A070101	Failure to Capture	04/04/2022	04/04/2022
				A0712	Pacing Problem	04/04/2022	04/04/2022
				A2303	Improper or Incorrect Procedure or Method	04/04/2022	04/04/2022
		F23	Unexpected Medical Intervention	A070101	Failure to Capture	04/04/2022	04/04/2022
				A0712	Pacing Problem	04/04/2022	04/04/2022
				A2303	Improper or Incorrect Procedure or Method	04/04/2022	04/04/2022
CT981	Single use	E0601	Arrhythmia	A070101	Failure to Capture	04/04/2022	04/04/2022
				A0712	Pacing Problem	04/04/2022	04/04/2022
				A2303	Improper or Incorrect Procedure or Method	04/04/2022	04/04/2022
		F23	Unexpected Medical Intervention	A070101	Failure to Capture	04/04/2022	04/04/2022
				A0712	Pacing Problem	04/04/2022	04/04/2022
				A2303	Improper or Incorrect Procedure or Method	04/04/2022	04/04/2022
CT983	Surgical invasive	E0601	Arrhythmia	A070101	Failure to Capture	04/04/2022	04/04/2022
				A0712	Pacing Problem	04/04/2022	04/04/2022
				A2303	Improper or Incorrect Procedure or Method	04/04/2022	04/04/2022
		F23	Unexpected Medical Intervention	A070101	Failure to Capture	04/04/2022	04/04/2022

									A0712	Pacing Problem	04/04/2022	04/04/2022
									A2303	Improper or Incorrect Procedure or Method	04/04/2022	04/04/2022
13704	2022	All other reportable incidents	37265	Dual-chamber implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E0601	Arrhythmia	A071302	Inappropriate/Inadequate Shock/Stimulation		
							F23	Unexpected Medical Intervention	A071302	Inappropriate/Inadequate Shock/Stimulation		
13705	2022	All other reportable incidents	35852	Single-chamber implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	11/05/2022	11/05/2022
							F1901	Additional Surgery	A070504	Premature Discharge of Battery	11/05/2022	11/05/2022
13706	2022	All other reportable incidents	47270	Cardiac resynchronization therapy implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A070910	Undersensing		
									A071302	Inappropriate/Inadequate Shock/Stimulation		
									A0908	Incorrect, Inadequate or Imprecise Result or Readings		
							F23	Unexpected Medical Intervention	A070910	Undersensing		
									A071302	Inappropriate/Inadequate Shock/Stimulation		
									A0908	Incorrect, Inadequate or Imprecise Result or Readings		
13707	2022	All other reportable incidents	35223	Endocardial/interventricular septal pacing lead	CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070909	Oversensing		
									A072202	Low Impedance		
							F1901	Additional Surgery	A070909	Oversensing		
									A072202	Low Impedance		
13708	2022	All other reportable incidents	47263	Cardiac resynchronization therapy implantable pacemaker	CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	01/06/2022	30/06/2022
							F1901	Additional Surgery	A070504	Premature Discharge of Battery	01/06/2022	30/06/2022
					CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	01/06/2022	30/06/2022
							F1901	Additional Surgery	A070504	Premature Discharge of Battery	01/06/2022	30/06/2022
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	01/06/2022	30/06/2022
							F1901	Additional Surgery	A070504	Premature Discharge of Battery	01/06/2022	30/06/2022
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	01/06/2022	30/06/2022
							F1901	Additional Surgery	A070504	Premature Discharge of Battery	01/06/2022	30/06/2022
13709	2022	All other reportable incidents	37265	Dual-chamber implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	24/06/2022	24/06/2022
							F1901	Additional Surgery	A070504	Premature Discharge of Battery	24/06/2022	24/06/2022
13710	2022	All other reportable incidents	47265	Dual-chamber implantable pacemaker, rate-responsive	CT141	Electrodes/Leads and associated devices	E0601	Arrhythmia	A070504	Premature Discharge of Battery	23/06/2022	23/06/2022
									A071205	Pacemaker Found in Back-Up Mode	23/06/2022	23/06/2022
									A090407	Therapy Delivered to Incorrect Body Area	23/06/2022	23/06/2022
							E2311	Discomfort	A070504	Premature Discharge of Battery	23/06/2022	23/06/2022
									A071205	Pacemaker Found in Back-Up Mode	23/06/2022	23/06/2022
									A090407	Therapy Delivered to Incorrect Body Area	23/06/2022	23/06/2022
							F1901	Additional Surgery	A070504	Premature Discharge of Battery	23/06/2022	23/06/2022
									A071205	Pacemaker Found in Back-Up Mode	23/06/2022	23/06/2022
									A090407	Therapy Delivered to Incorrect Body Area	23/06/2022	23/06/2022
					CT145	Cardiovascular devices	E0601	Arrhythmia	A070504	Premature Discharge of Battery	23/06/2022	23/06/2022
									A071205	Pacemaker Found in Back-Up Mode	23/06/2022	23/06/2022
									A090407	Therapy Delivered to Incorrect Body Area	23/06/2022	23/06/2022
							E2311	Discomfort	A070504	Premature Discharge of Battery	23/06/2022	23/06/2022
									A071205	Pacemaker Found in Back-Up Mode	23/06/2022	23/06/2022
									A090407	Therapy Delivered to Incorrect Body Area	23/06/2022	23/06/2022
							F1901	Additional Surgery	A070504	Premature Discharge of Battery	23/06/2022	23/06/2022
									A071205	Pacemaker Found in Back-Up Mode	23/06/2022	23/06/2022
									A090407	Therapy Delivered to Incorrect Body Area	23/06/2022	23/06/2022
					CT1684	Cardiac pacemakers and associated devices	E0601	Arrhythmia	A070504	Premature Discharge of Battery	23/06/2022	23/06/2022
									A071205	Pacemaker Found in Back-Up Mode	23/06/2022	23/06/2022
									A090407	Therapy Delivered to Incorrect Body Area	23/06/2022	23/06/2022
							E2311	Discomfort	A070504	Premature Discharge of Battery	23/06/2022	23/06/2022
									A071205	Pacemaker Found in Back-Up Mode	23/06/2022	23/06/2022
									A090407	Therapy Delivered to Incorrect Body Area	23/06/2022	23/06/2022
							F1901	Additional Surgery	A070504	Premature Discharge of Battery	23/06/2022	23/06/2022
									A071205	Pacemaker Found in Back-Up Mode	23/06/2022	23/06/2022
									A090407	Therapy Delivered to Incorrect Body Area	23/06/2022	23/06/2022
					CT1685	Cardiac pacemakers	E0601	Arrhythmia	A070504	Premature Discharge of Battery	23/06/2022	23/06/2022
									A071205	Pacemaker Found in Back-Up Mode	23/06/2022	23/06/2022
									A090407	Therapy Delivered to Incorrect Body Area	23/06/2022	23/06/2022
							E2311	Discomfort	A070504	Premature Discharge of Battery	23/06/2022	23/06/2022
									A071205	Pacemaker Found in Back-Up Mode	23/06/2022	23/06/2022
									A090407	Therapy Delivered to Incorrect Body Area	23/06/2022	23/06/2022
							F1901	Additional Surgery	A070504	Premature Discharge of Battery	23/06/2022	23/06/2022
									A071205	Pacemaker Found in Back-Up Mode	23/06/2022	23/06/2022
									A090407	Therapy Delivered to Incorrect Body Area	23/06/2022	23/06/2022
					CT233	Surgical	E0601	Arrhythmia	A070504	Premature Discharge of Battery	23/06/2022	23/06/2022
									A071205	Pacemaker Found in Back-Up Mode	23/06/2022	23/06/2022
									A090407	Therapy Delivered to Incorrect Body Area	23/06/2022	23/06/2022
							E2311	Discomfort	A070504	Premature Discharge of Battery	23/06/2022	23/06/2022
									A071205	Pacemaker Found in Back-Up Mode	23/06/2022	23/06/2022
									A090407	Therapy Delivered to Incorrect Body Area	23/06/2022	23/06/2022
							F1901	Additional Surgery	A070504	Premature Discharge of Battery	23/06/2022	23/06/2022
									A071205	Pacemaker Found in Back-Up Mode	23/06/2022	23/06/2022
									A090407	Therapy Delivered to Incorrect Body Area	23/06/2022	23/06/2022
					CT2406	Implantable	E0601	Arrhythmia	A070504	Premature Discharge of Battery	23/06/2022	23/06/2022
									A071205	Pacemaker Found in Back-Up Mode	23/06/2022	23/06/2022
									A090407	Therapy Delivered to Incorrect Body Area	23/06/2022	23/06/2022
							E2311	Discomfort	A070504	Premature Discharge of Battery	23/06/2022	23/06/2022
									A071205	Pacemaker Found in Back-Up Mode	23/06/2022	23/06/2022
									A090407	Therapy Delivered to Incorrect Body Area	23/06/2022	23/06/2022
							F1901	Additional Surgery	A070504	Premature Discharge of Battery	23/06/2022	23/06/2022
									A071205	Pacemaker Found in Back-Up Mode	23/06/2022	23/06/2022
									A090407	Therapy Delivered to Incorrect Body Area	23/06/2022	23/06/2022
					CT2407	Active implantable	E0601	Arrhythmia	A070504	Premature Discharge of Battery	23/06/2022	23/06/2022
									A071205	Pacemaker Found in Back-Up Mode	23/06/2022	23/06/2022
									A090407	Therapy Delivered to Incorrect Body Area	23/06/2022	23/06/2022
							E2311	Discomfort	A070504	Premature Discharge of Battery	23/06/2022	23/06/2022
									A071205	Pacemaker Found in Back-Up Mode	23/06/2022	23/06/2022
									A090407	Therapy Delivered to Incorrect Body Area	23/06/2022	23/06/2022

							F1901	Additional Surgery	A070504	Premature Discharge of Battery	23/06/2022	23/06/2022
									A071205	Pacemaker Found in Back-Up Mode	23/06/2022	23/06/2022
									A090407	Therapy Delivered to Incorrect Body Area	23/06/2022	23/06/2022
					CT257	Implantable cardiac pacemakers	E0601	Arrhythmia	A070504	Premature Discharge of Battery	23/06/2022	23/06/2022
									A071205	Pacemaker Found in Back-Up Mode	23/06/2022	23/06/2022
									A090407	Therapy Delivered to Incorrect Body Area	23/06/2022	23/06/2022
							E2311	Discomfort	A070504	Premature Discharge of Battery	23/06/2022	23/06/2022
									A071205	Pacemaker Found in Back-Up Mode	23/06/2022	23/06/2022
									A090407	Therapy Delivered to Incorrect Body Area	23/06/2022	23/06/2022
							F1901	Additional Surgery	A070504	Premature Discharge of Battery	23/06/2022	23/06/2022
									A071205	Pacemaker Found in Back-Up Mode	23/06/2022	23/06/2022
									A090407	Therapy Delivered to Incorrect Body Area	23/06/2022	23/06/2022
					CT2727	Active implant devices/ancillaries	E0601	Arrhythmia	A070504	Premature Discharge of Battery	23/06/2022	23/06/2022
									A071205	Pacemaker Found in Back-Up Mode	23/06/2022	23/06/2022
									A090407	Therapy Delivered to Incorrect Body Area	23/06/2022	23/06/2022
							E2311	Discomfort	A070504	Premature Discharge of Battery	23/06/2022	23/06/2022
									A071205	Pacemaker Found in Back-Up Mode	23/06/2022	23/06/2022
									A090407	Therapy Delivered to Incorrect Body Area	23/06/2022	23/06/2022
							F1901	Additional Surgery	A070504	Premature Discharge of Battery	23/06/2022	23/06/2022
									A071205	Pacemaker Found in Back-Up Mode	23/06/2022	23/06/2022
									A090407	Therapy Delivered to Incorrect Body Area	23/06/2022	23/06/2022
					CT2750	Implantable cardiac electrical stimulators and associated devices	E0601	Arrhythmia	A070504	Premature Discharge of Battery	23/06/2022	23/06/2022
									A071205	Pacemaker Found in Back-Up Mode	23/06/2022	23/06/2022
									A090407	Therapy Delivered to Incorrect Body Area	23/06/2022	23/06/2022
							E2311	Discomfort	A070504	Premature Discharge of Battery	23/06/2022	23/06/2022
									A071205	Pacemaker Found in Back-Up Mode	23/06/2022	23/06/2022
									A090407	Therapy Delivered to Incorrect Body Area	23/06/2022	23/06/2022
							F1901	Additional Surgery	A070504	Premature Discharge of Battery	23/06/2022	23/06/2022
									A071205	Pacemaker Found in Back-Up Mode	23/06/2022	23/06/2022
									A090407	Therapy Delivered to Incorrect Body Area	23/06/2022	23/06/2022
					CT2866	Implantable cardiac electrical stimulators	E0601	Arrhythmia	A070504	Premature Discharge of Battery	23/06/2022	23/06/2022
									A071205	Pacemaker Found in Back-Up Mode	23/06/2022	23/06/2022
									A090407	Therapy Delivered to Incorrect Body Area	23/06/2022	23/06/2022
							E2311	Discomfort	A070504	Premature Discharge of Battery	23/06/2022	23/06/2022
									A071205	Pacemaker Found in Back-Up Mode	23/06/2022	23/06/2022
									A090407	Therapy Delivered to Incorrect Body Area	23/06/2022	23/06/2022
							F1901	Additional Surgery	A070504	Premature Discharge of Battery	23/06/2022	23/06/2022
									A071205	Pacemaker Found in Back-Up Mode	23/06/2022	23/06/2022
									A090407	Therapy Delivered to Incorrect Body Area	23/06/2022	23/06/2022
					CT2985	Active	E0601	Arrhythmia	A070504	Premature Discharge of Battery	23/06/2022	23/06/2022
									A071205	Pacemaker Found in Back-Up Mode	23/06/2022	23/06/2022
									A090407	Therapy Delivered to Incorrect Body Area	23/06/2022	23/06/2022
							E2311	Discomfort	A070504	Premature Discharge of Battery	23/06/2022	23/06/2022
									A071205	Pacemaker Found in Back-Up Mode	23/06/2022	23/06/2022
									A090407	Therapy Delivered to Incorrect Body Area	23/06/2022	23/06/2022
							F1901	Additional Surgery	A070504	Premature Discharge of Battery	23/06/2022	23/06/2022
									A071205	Pacemaker Found in Back-Up Mode	23/06/2022	23/06/2022
									A090407	Therapy Delivered to Incorrect Body Area	23/06/2022	23/06/2022
					CT3001	Implantable metal	E0601	Arrhythmia	A070504	Premature Discharge of Battery	23/06/2022	23/06/2022
									A071205	Pacemaker Found in Back-Up Mode	23/06/2022	23/06/2022
									A090407	Therapy Delivered to Incorrect Body Area	23/06/2022	23/06/2022
							E2311	Discomfort	A070504	Premature Discharge of Battery	23/06/2022	23/06/2022
									A071205	Pacemaker Found in Back-Up Mode	23/06/2022	23/06/2022
									A090407	Therapy Delivered to Incorrect Body Area	23/06/2022	23/06/2022
							F1901	Additional Surgery	A070504	Premature Discharge of Battery	23/06/2022	23/06/2022
									A071205	Pacemaker Found in Back-Up Mode	23/06/2022	23/06/2022
									A090407	Therapy Delivered to Incorrect Body Area	23/06/2022	23/06/2022
					CT3002	Implantable synthetic polymer	E0601	Arrhythmia	A070504	Premature Discharge of Battery	23/06/2022	23/06/2022
									A071205	Pacemaker Found in Back-Up Mode	23/06/2022	23/06/2022
									A090407	Therapy Delivered to Incorrect Body Area	23/06/2022	23/06/2022
							E2311	Discomfort	A070504	Premature Discharge of Battery	23/06/2022	23/06/2022
									A071205	Pacemaker Found in Back-Up Mode	23/06/2022	23/06/2022
									A090407	Therapy Delivered to Incorrect Body Area	23/06/2022	23/06/2022
							F1901	Additional Surgery	A070504	Premature Discharge of Battery	23/06/2022	23/06/2022
									A071205	Pacemaker Found in Back-Up Mode	23/06/2022	23/06/2022
									A090407	Therapy Delivered to Incorrect Body Area	23/06/2022	23/06/2022
					CT3006	Implantable inorganic material	E0601	Arrhythmia	A070504	Premature Discharge of Battery	23/06/2022	23/06/2022
									A071205	Pacemaker Found in Back-Up Mode	23/06/2022	23/06/2022
									A090407	Therapy Delivered to Incorrect Body Area	23/06/2022	23/06/2022
							E2311	Discomfort	A070504	Premature Discharge of Battery	23/06/2022	23/06/2022
									A071205	Pacemaker Found in Back-Up Mode	23/06/2022	23/06/2022
									A090407	Therapy Delivered to Incorrect Body Area	23/06/2022	23/06/2022
							F1901	Additional Surgery	A070504	Premature Discharge of Battery	23/06/2022	23/06/2022
									A071205	Pacemaker Found in Back-Up Mode	23/06/2022	23/06/2022
									A090407	Therapy Delivered to Incorrect Body Area	23/06/2022	23/06/2022
					CT314	Electrically powered	E0601	Arrhythmia	A070504	Premature Discharge of Battery	23/06/2022	23/06/2022
									A071205	Pacemaker Found in Back-Up Mode	23/06/2022	23/06/2022
									A090407	Therapy Delivered to Incorrect Body Area	23/06/2022	23/06/2022
							E2311	Discomfort	A070504	Premature Discharge of Battery	23/06/2022	23/06/2022
									A071205	Pacemaker Found in Back-Up Mode	23/06/2022	23/06/2022
									A090407	Therapy Delivered to Incorrect Body Area	23/06/2022	23/06/2022
							F1901	Additional Surgery	A070504	Premature Discharge of Battery	23/06/2022	23/06/2022
									A071205	Pacemaker Found in Back-Up Mode	23/06/2022	23/06/2022
									A090407	Therapy Delivered to Incorrect Body Area	23/06/2022	23/06/2022
					CT321	Long-term surgical invasive	E0601	Arrhythmia	A070504	Premature Discharge of Battery	23/06/2022	23/06/2022
									A071205	Pacemaker Found in Back-Up Mode	23/06/2022	23/06/2022
									A090407	Therapy Delivered to Incorrect Body Area	23/06/2022	23/06/2022
							E2311	Discomfort	A070504	Premature Discharge of Battery	23/06/2022	23/06/2022

									A071205	Pacemaker Found in Back-Up Mode	23/06/2022	23/06/2022
									A090407	Therapy Delivered to Incorrect Body Area	23/06/2022	23/06/2022
							F1901	Additional Surgery	A070504	Premature Discharge of Battery	23/06/2022	23/06/2022
									A071205	Pacemaker Found in Back-Up Mode	23/06/2022	23/06/2022
									A090407	Therapy Delivered to Incorrect Body Area	23/06/2022	23/06/2022
					CT449	Generators and associated devices	E0601	Arrhythmia	A070504	Premature Discharge of Battery	23/06/2022	23/06/2022
									A071205	Pacemaker Found in Back-Up Mode	23/06/2022	23/06/2022
									A090407	Therapy Delivered to Incorrect Body Area	23/06/2022	23/06/2022
							E2311	Discomfort	A070504	Premature Discharge of Battery	23/06/2022	23/06/2022
									A071205	Pacemaker Found in Back-Up Mode	23/06/2022	23/06/2022
									A090407	Therapy Delivered to Incorrect Body Area	23/06/2022	23/06/2022
							F1901	Additional Surgery	A070504	Premature Discharge of Battery	23/06/2022	23/06/2022
									A071205	Pacemaker Found in Back-Up Mode	23/06/2022	23/06/2022
									A090407	Therapy Delivered to Incorrect Body Area	23/06/2022	23/06/2022
					CT981	Single use	E0601	Arrhythmia	A070504	Premature Discharge of Battery	23/06/2022	23/06/2022
									A071205	Pacemaker Found in Back-Up Mode	23/06/2022	23/06/2022
									A090407	Therapy Delivered to Incorrect Body Area	23/06/2022	23/06/2022
							E2311	Discomfort	A070504	Premature Discharge of Battery	23/06/2022	23/06/2022
									A071205	Pacemaker Found in Back-Up Mode	23/06/2022	23/06/2022
									A090407	Therapy Delivered to Incorrect Body Area	23/06/2022	23/06/2022
							F1901	Additional Surgery	A070504	Premature Discharge of Battery	23/06/2022	23/06/2022
									A071205	Pacemaker Found in Back-Up Mode	23/06/2022	23/06/2022
									A090407	Therapy Delivered to Incorrect Body Area	23/06/2022	23/06/2022
					CT983	Surgical invasive	E0601	Arrhythmia	A070504	Premature Discharge of Battery	23/06/2022	23/06/2022
									A071205	Pacemaker Found in Back-Up Mode	23/06/2022	23/06/2022
									A090407	Therapy Delivered to Incorrect Body Area	23/06/2022	23/06/2022
							E2311	Discomfort	A070504	Premature Discharge of Battery	23/06/2022	23/06/2022
									A071205	Pacemaker Found in Back-Up Mode	23/06/2022	23/06/2022
									A090407	Therapy Delivered to Incorrect Body Area	23/06/2022	23/06/2022
							F1901	Additional Surgery	A070504	Premature Discharge of Battery	23/06/2022	23/06/2022
									A071205	Pacemaker Found in Back-Up Mode	23/06/2022	23/06/2022
									A090407	Therapy Delivered to Incorrect Body Area	23/06/2022	23/06/2022
13711	2022	All other reportable incidents	47265	Dual-chamber implantable pacemaker, rate-responsive	CT141	Electrodes/Leads and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	18/09/2022	18/09/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	18/09/2022	18/09/2022
					CT145	Cardiovascular devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	18/09/2022	18/09/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	18/09/2022	18/09/2022
					CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	18/09/2022	18/09/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	18/09/2022	18/09/2022
					CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	18/09/2022	18/09/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	18/09/2022	18/09/2022
					CT233	Surgical	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	18/09/2022	18/09/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	18/09/2022	18/09/2022
					CT2406	Implantable	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	18/09/2022	18/09/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	18/09/2022	18/09/2022
					CT2407	Active implantable	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	18/09/2022	18/09/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	18/09/2022	18/09/2022
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	18/09/2022	18/09/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	18/09/2022	18/09/2022
					CT2727	Active implant devices/ancillaries	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	18/09/2022	18/09/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	18/09/2022	18/09/2022
					CT2750	Implantable cardiac electrical stimulators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	18/09/2022	18/09/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	18/09/2022	18/09/2022
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	18/09/2022	18/09/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	18/09/2022	18/09/2022
					CT2985	Active	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	18/09/2022	18/09/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	18/09/2022	18/09/2022
					CT3001	Implantable metal	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	18/09/2022	18/09/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	18/09/2022	18/09/2022
					CT3002	Implantable synthetic polymer	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	18/09/2022	18/09/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	18/09/2022	18/09/2022
					CT3006	Implantable inorganic material	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	18/09/2022	18/09/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	18/09/2022	18/09/2022
					CT314	Electrically powered	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	18/09/2022	18/09/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	18/09/2022	18/09/2022
					CT321	Long-term surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	18/09/2022	18/09/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	18/09/2022	18/09/2022
					CT449	Generators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	18/09/2022	18/09/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	18/09/2022	18/09/2022
					CT981	Single use	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	18/09/2022	18/09/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	18/09/2022	18/09/2022
					CT983	Surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	18/09/2022	18/09/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	18/09/2022	18/09/2022
13712	2022	All other reportable incidents	47265	Dual-chamber implantable pacemaker, rate-responsive	CT141	Electrodes/Leads and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	02/09/2022	02/09/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	02/09/2022	02/09/2022
					CT145	Cardiovascular devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	02/09/2022	02/09/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	02/09/2022	02/09/2022
					CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	02/09/2022	02/09/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	02/09/2022	02/09/2022
					CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	02/09/2022	02/09/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	02/09/2022	02/09/2022
					CT233	Surgical	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	02/09/2022	02/09/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	02/09/2022	02/09/2022
					CT2406	Implantable	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	02/09/2022	02/09/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	02/09/2022	02/09/2022
					CT2407	Active implantable	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	02/09/2022	02/09/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	02/09/2022	02/09/2022
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	02/09/2022	02/09/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	02/09/2022	02/09/2022

					CT2727	Active implant devices/ancillaries	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	02/09/2022	02/09/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	02/09/2022	02/09/2022
					CT2750	Implantable cardiac electrical stimulators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	02/09/2022	02/09/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	02/09/2022	02/09/2022
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	02/09/2022	02/09/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	02/09/2022	02/09/2022
					CT2985	Active	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	02/09/2022	02/09/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	02/09/2022	02/09/2022
					CT3001	Implantable metal	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	02/09/2022	02/09/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	02/09/2022	02/09/2022
					CT3002	Implantable synthetic polymer	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	02/09/2022	02/09/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	02/09/2022	02/09/2022
					CT3006	Implantable inorganic material	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	02/09/2022	02/09/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	02/09/2022	02/09/2022
					CT314	Electrically powered	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	02/09/2022	02/09/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	02/09/2022	02/09/2022
					CT321	Long-term surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	02/09/2022	02/09/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	02/09/2022	02/09/2022
					CT449	Generators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	02/09/2022	02/09/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	02/09/2022	02/09/2022
					CT981	Single use	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	02/09/2022	02/09/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	02/09/2022	02/09/2022
					CT983	Surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	02/09/2022	02/09/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	02/09/2022	02/09/2022
13713	2022	All other reportable incidents	35223	Endocardial/interventricular septal pacing lead	CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A090801	Signal Artifact/Noise		
							F1901	Additional Surgery	A090801	Signal Artifact/Noise		
13714	2022	All other reportable incidents	35223	Endocardial/interventricular septal pacing lead	CT1684	Cardiac pacemakers and associated devices	E0717	Dyspnea	A040101	Fracture		
									A070909	Oversensing		
							F11	Non-Serious Injury/ Illness/ Impairment	A040101	Fracture		
									A070909	Oversensing		
13715	2022	All other reportable incidents	47265	Dual-chamber implantable pacemaker, rate-responsive	CT141	Electrodes/Leads and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	08/10/2022	08/10/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	08/10/2022	08/10/2022
					CT145	Cardiovascular devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	08/10/2022	08/10/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	08/10/2022	08/10/2022
					CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	08/10/2022	08/10/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	08/10/2022	08/10/2022
					CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	08/10/2022	08/10/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	08/10/2022	08/10/2022
					CT233	Surgical	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	08/10/2022	08/10/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	08/10/2022	08/10/2022
					CT2406	Implantable	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	08/10/2022	08/10/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	08/10/2022	08/10/2022
					CT2407	Active implantable	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	08/10/2022	08/10/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	08/10/2022	08/10/2022
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	08/10/2022	08/10/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	08/10/2022	08/10/2022
					CT2727	Active implant devices/ancillaries	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	08/10/2022	08/10/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	08/10/2022	08/10/2022
					CT2750	Implantable cardiac electrical stimulators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	08/10/2022	08/10/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	08/10/2022	08/10/2022
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	08/10/2022	08/10/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	08/10/2022	08/10/2022
					CT2985	Active	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	08/10/2022	08/10/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	08/10/2022	08/10/2022
					CT3001	Implantable metal	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	08/10/2022	08/10/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	08/10/2022	08/10/2022
					CT3002	Implantable synthetic polymer	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	08/10/2022	08/10/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	08/10/2022	08/10/2022
					CT3006	Implantable inorganic material	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	08/10/2022	08/10/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	08/10/2022	08/10/2022
					CT314	Electrically powered	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	08/10/2022	08/10/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	08/10/2022	08/10/2022
					CT321	Long-term surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	08/10/2022	08/10/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	08/10/2022	08/10/2022
					CT449	Generators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	08/10/2022	08/10/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	08/10/2022	08/10/2022
					CT981	Single use	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	08/10/2022	08/10/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	08/10/2022	08/10/2022
					CT983	Surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	08/10/2022	08/10/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	08/10/2022	08/10/2022
13716	2022	All other reportable incidents	47265	Dual-chamber implantable pacemaker, rate-responsive	CT141	Electrodes/Leads and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	08/10/2022	08/10/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	08/10/2022	08/10/2022
					CT145	Cardiovascular devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	08/10/2022	08/10/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	08/10/2022	08/10/2022
					CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	08/10/2022	08/10/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	08/10/2022	08/10/2022
					CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	08/10/2022	08/10/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	08/10/2022	08/10/2022
					CT233	Surgical	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	08/10/2022	08/10/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	08/10/2022	08/10/2022
					CT2406	Implantable	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	08/10/2022	08/10/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	08/10/2022	08/10/2022
					CT2407	Active implantable	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	08/10/2022	08/10/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	08/10/2022	08/10/2022
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	08/10/2022	08/10/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	08/10/2022	08/10/2022
					CT2727	Active implant devices/ancillaries	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	08/10/2022	08/10/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	08/10/2022	08/10/2022

					CT2750	Implantable cardiac electrical stimulators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	08/10/2022	08/10/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	08/10/2022	08/10/2022
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	08/10/2022	08/10/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	08/10/2022	08/10/2022
					CT2985	Active	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	08/10/2022	08/10/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	08/10/2022	08/10/2022
					CT3001	Implantable metal	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	08/10/2022	08/10/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	08/10/2022	08/10/2022
					CT3002	Implantable synthetic polymer	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	08/10/2022	08/10/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	08/10/2022	08/10/2022
					CT3006	Implantable inorganic material	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	08/10/2022	08/10/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	08/10/2022	08/10/2022
					CT314	Electrically powered	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	08/10/2022	08/10/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	08/10/2022	08/10/2022
					CT321	Long-term surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	08/10/2022	08/10/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	08/10/2022	08/10/2022
					CT449	Generators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	08/10/2022	08/10/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	08/10/2022	08/10/2022
					CT981	Single use	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	08/10/2022	08/10/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	08/10/2022	08/10/2022
					CT983	Surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	08/10/2022	08/10/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	08/10/2022	08/10/2022
13717	2022	All other reportable incidents	47265	Dual-chamber implantable pacemaker, rate-responsive	CT141	Electrodes/Leads and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	14/10/2022	14/10/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	14/10/2022	14/10/2022
					CT145	Cardiovascular devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	14/10/2022	14/10/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	14/10/2022	14/10/2022
					CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	14/10/2022	14/10/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	14/10/2022	14/10/2022
					CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	14/10/2022	14/10/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	14/10/2022	14/10/2022
					CT233	Surgical	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	14/10/2022	14/10/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	14/10/2022	14/10/2022
					CT2406	Implantable	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	14/10/2022	14/10/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	14/10/2022	14/10/2022
					CT2407	Active implantable	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	14/10/2022	14/10/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	14/10/2022	14/10/2022
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	14/10/2022	14/10/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	14/10/2022	14/10/2022
					CT2727	Active implant devices/ancillaries	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	14/10/2022	14/10/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	14/10/2022	14/10/2022
					CT2750	Implantable cardiac electrical stimulators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	14/10/2022	14/10/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	14/10/2022	14/10/2022
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	14/10/2022	14/10/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	14/10/2022	14/10/2022
					CT2985	Active	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	14/10/2022	14/10/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	14/10/2022	14/10/2022
					CT3001	Implantable metal	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	14/10/2022	14/10/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	14/10/2022	14/10/2022
					CT3002	Implantable synthetic polymer	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	14/10/2022	14/10/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	14/10/2022	14/10/2022
					CT3006	Implantable inorganic material	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	14/10/2022	14/10/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	14/10/2022	14/10/2022
					CT314	Electrically powered	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	14/10/2022	14/10/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	14/10/2022	14/10/2022
					CT321	Long-term surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	14/10/2022	14/10/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	14/10/2022	14/10/2022
					CT449	Generators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	14/10/2022	14/10/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	14/10/2022	14/10/2022
					CT981	Single use	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	14/10/2022	14/10/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	14/10/2022	14/10/2022
					CT983	Surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	14/10/2022	14/10/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	14/10/2022	14/10/2022
13718	2022	All other reportable incidents	35223	Endocardial/interventricular septal pacing lead	CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A0401	Break	30/06/2022	30/06/2022
									A070909	Oversensing	30/06/2022	30/06/2022
									A0712	Pacing Problem	30/06/2022	30/06/2022
							F1901	Additional Surgery	A0401	Break	30/06/2022	30/06/2022
									A070909	Oversensing	30/06/2022	30/06/2022
									A0712	Pacing Problem	30/06/2022	30/06/2022
13719	2022	All other reportable incidents	47265	Dual-chamber implantable pacemaker, rate-responsive	CT141	Electrodes/Leads and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT145	Cardiovascular devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT233	Surgical	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT2406	Implantable	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT2407	Active implantable	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT2727	Active implant devices/ancillaries	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT2750	Implantable cardiac electrical stimulators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		

					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT2985	Active	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT3001	Implantable metal	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT3002	Implantable synthetic polymer	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT3006	Implantable inorganic material	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT314	Electrically powered	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT321	Long-term surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT449	Generators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT981	Single use	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT983	Surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
13720	2022	All other reportable incidents	47265	Dual-chamber implantable pacemaker, rate-responsive	CT141	Electrodes/Leads and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	17/10/2022	17/10/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	17/10/2022	17/10/2022
					CT145	Cardiovascular devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	17/10/2022	17/10/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	17/10/2022	17/10/2022
					CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	17/10/2022	17/10/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	17/10/2022	17/10/2022
					CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	17/10/2022	17/10/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	17/10/2022	17/10/2022
					CT233	Surgical	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	17/10/2022	17/10/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	17/10/2022	17/10/2022
					CT2406	Implantable	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	17/10/2022	17/10/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	17/10/2022	17/10/2022
					CT2407	Active implantable	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	17/10/2022	17/10/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	17/10/2022	17/10/2022
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	17/10/2022	17/10/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	17/10/2022	17/10/2022
					CT2727	Active implant devices/ancillaries	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	17/10/2022	17/10/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	17/10/2022	17/10/2022
					CT2750	Implantable cardiac electrical stimulators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	17/10/2022	17/10/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	17/10/2022	17/10/2022
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	17/10/2022	17/10/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	17/10/2022	17/10/2022
					CT2985	Active	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	17/10/2022	17/10/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	17/10/2022	17/10/2022
					CT3001	Implantable metal	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	17/10/2022	17/10/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	17/10/2022	17/10/2022
					CT3002	Implantable synthetic polymer	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	17/10/2022	17/10/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	17/10/2022	17/10/2022
					CT3006	Implantable inorganic material	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	17/10/2022	17/10/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	17/10/2022	17/10/2022
					CT314	Electrically powered	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	17/10/2022	17/10/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	17/10/2022	17/10/2022
					CT321	Long-term surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	17/10/2022	17/10/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	17/10/2022	17/10/2022
					CT449	Generators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	17/10/2022	17/10/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	17/10/2022	17/10/2022
					CT981	Single use	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	17/10/2022	17/10/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	17/10/2022	17/10/2022
					CT983	Surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	17/10/2022	17/10/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	17/10/2022	17/10/2022
13721	2022	All other reportable incidents	47265	Dual-chamber implantable pacemaker, rate-responsive	CT141	Electrodes/Leads and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT145	Cardiovascular devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT233	Surgical	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT2406	Implantable	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT2407	Active implantable	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT2727	Active implant devices/ancillaries	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT2750	Implantable cardiac electrical stimulators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT2985	Active	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT3001	Implantable metal	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT3002	Implantable synthetic polymer	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		

					CT3006	Implantable inorganic material	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT314	Electrically powered	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT321	Long-term surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT449	Generators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT981	Single use	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT983	Surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
13722	2022	All other reportable incidents	47265	Dual-chamber implantable pacemaker, rate-responsive	CT141	Electrodes/Leads and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	27/10/2022	27/10/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	27/10/2022	27/10/2022
					CT145	Cardiovascular devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	27/10/2022	27/10/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	27/10/2022	27/10/2022
					CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	27/10/2022	27/10/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	27/10/2022	27/10/2022
					CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	27/10/2022	27/10/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	27/10/2022	27/10/2022
					CT233	Surgical	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	27/10/2022	27/10/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	27/10/2022	27/10/2022
					CT2406	Implantable	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	27/10/2022	27/10/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	27/10/2022	27/10/2022
					CT2407	Active implantable	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	27/10/2022	27/10/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	27/10/2022	27/10/2022
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	27/10/2022	27/10/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	27/10/2022	27/10/2022
					CT2727	Active implant devices/ancillaries	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	27/10/2022	27/10/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	27/10/2022	27/10/2022
					CT2750	Implantable cardiac electrical stimulators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	27/10/2022	27/10/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	27/10/2022	27/10/2022
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	27/10/2022	27/10/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	27/10/2022	27/10/2022
					CT2985	Active	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	27/10/2022	27/10/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	27/10/2022	27/10/2022
					CT3001	Implantable metal	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	27/10/2022	27/10/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	27/10/2022	27/10/2022
					CT3002	Implantable synthetic polymer	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	27/10/2022	27/10/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	27/10/2022	27/10/2022
					CT3006	Implantable inorganic material	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	27/10/2022	27/10/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	27/10/2022	27/10/2022
					CT314	Electrically powered	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	27/10/2022	27/10/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	27/10/2022	27/10/2022
					CT321	Long-term surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	27/10/2022	27/10/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	27/10/2022	27/10/2022
					CT449	Generators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	27/10/2022	27/10/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	27/10/2022	27/10/2022
					CT981	Single use	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	27/10/2022	27/10/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	27/10/2022	27/10/2022
					CT983	Surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	27/10/2022	27/10/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	27/10/2022	27/10/2022
13723	2022	All other reportable incidents	47265	Dual-chamber implantable pacemaker, rate-responsive	CT141	Electrodes/Leads and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A0711	Interrogation Problem		
									A0712	Pacing Problem		
							F26	No Health Consequences or Impact	A0711	Interrogation Problem		
									A0712	Pacing Problem		
					CT145	Cardiovascular devices	E2403	No Clinical Signs, Symptoms or Conditions	A0711	Interrogation Problem		
									A0712	Pacing Problem		
							F26	No Health Consequences or Impact	A0711	Interrogation Problem		
									A0712	Pacing Problem		
					CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A0711	Interrogation Problem		
									A0712	Pacing Problem		
							F26	No Health Consequences or Impact	A0711	Interrogation Problem		
									A0712	Pacing Problem		
					CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A0711	Interrogation Problem		
									A0712	Pacing Problem		
							F26	No Health Consequences or Impact	A0711	Interrogation Problem		
									A0712	Pacing Problem		
					CT233	Surgical	E2403	No Clinical Signs, Symptoms or Conditions	A0711	Interrogation Problem		
									A0712	Pacing Problem		
							F26	No Health Consequences or Impact	A0711	Interrogation Problem		
									A0712	Pacing Problem		
					CT2406	Implantable	E2403	No Clinical Signs, Symptoms or Conditions	A0711	Interrogation Problem		
									A0712	Pacing Problem		
							F26	No Health Consequences or Impact	A0711	Interrogation Problem		
									A0712	Pacing Problem		
					CT2407	Active implantable	E2403	No Clinical Signs, Symptoms or Conditions	A0711	Interrogation Problem		
									A0712	Pacing Problem		
							F26	No Health Consequences or Impact	A0711	Interrogation Problem		
									A0712	Pacing Problem		
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A0711	Interrogation Problem		
									A0712	Pacing Problem		
							F26	No Health Consequences or Impact	A0711	Interrogation Problem		
									A0712	Pacing Problem		
					CT2727	Active implant devices/ancillaries	E2403	No Clinical Signs, Symptoms or Conditions	A0711	Interrogation Problem		
									A0712	Pacing Problem		
							F26	No Health Consequences or Impact	A0711	Interrogation Problem		
									A0712	Pacing Problem		

					CT2750	Implantable cardiac electrical stimulators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A0711	Interrogation Problem		
									A0712	Pacing Problem		
							F26	No Health Consequences or Impact	A0711	Interrogation Problem		
									A0712	Pacing Problem		
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A0711	Interrogation Problem		
									A0712	Pacing Problem		
							F26	No Health Consequences or Impact	A0711	Interrogation Problem		
									A0712	Pacing Problem		
					CT2985	Active	E2403	No Clinical Signs, Symptoms or Conditions	A0711	Interrogation Problem		
									A0712	Pacing Problem		
							F26	No Health Consequences or Impact	A0711	Interrogation Problem		
									A0712	Pacing Problem		
					CT3001	Implantable metal	E2403	No Clinical Signs, Symptoms or Conditions	A0711	Interrogation Problem		
									A0712	Pacing Problem		
							F26	No Health Consequences or Impact	A0711	Interrogation Problem		
									A0712	Pacing Problem		
					CT3002	Implantable synthetic polymer	E2403	No Clinical Signs, Symptoms or Conditions	A0711	Interrogation Problem		
									A0712	Pacing Problem		
							F26	No Health Consequences or Impact	A0711	Interrogation Problem		
									A0712	Pacing Problem		
					CT3006	Implantable inorganic material	E2403	No Clinical Signs, Symptoms or Conditions	A0711	Interrogation Problem		
									A0712	Pacing Problem		
							F26	No Health Consequences or Impact	A0711	Interrogation Problem		
									A0712	Pacing Problem		
					CT314	Electrically powered	E2403	No Clinical Signs, Symptoms or Conditions	A0711	Interrogation Problem		
									A0712	Pacing Problem		
							F26	No Health Consequences or Impact	A0711	Interrogation Problem		
									A0712	Pacing Problem		
					CT321	Long-term surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A0711	Interrogation Problem		
									A0712	Pacing Problem		
							F26	No Health Consequences or Impact	A0711	Interrogation Problem		
									A0712	Pacing Problem		
					CT449	Generators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A0711	Interrogation Problem		
									A0712	Pacing Problem		
							F26	No Health Consequences or Impact	A0711	Interrogation Problem		
									A0712	Pacing Problem		
					CT981	Single use	E2403	No Clinical Signs, Symptoms or Conditions	A0711	Interrogation Problem		
									A0712	Pacing Problem		
							F26	No Health Consequences or Impact	A0711	Interrogation Problem		
									A0712	Pacing Problem		
					CT983	Surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A0711	Interrogation Problem		
									A0712	Pacing Problem		
							F26	No Health Consequences or Impact	A0711	Interrogation Problem		
									A0712	Pacing Problem		
13724	2022	All other reportable incidents	35223	Endocardial/interventricular septal pacing lead	CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070101	Failure to Capture		
							F23	Unexpected Medical Intervention	A070101	Failure to Capture		
13725	2022	All other reportable incidents	47265	Dual-chamber implantable pacemaker, rate-responsive	CT141	Electrodes/Leads and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem		
							F26	No Health Consequences or Impact	A24	No Apparent Device Problem		
					CT145	Cardiovascular devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem		
							F26	No Health Consequences or Impact	A24	No Apparent Device Problem		
					CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem		
							F26	No Health Consequences or Impact	A24	No Apparent Device Problem		
					CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem		
							F26	No Health Consequences or Impact	A24	No Apparent Device Problem		
					CT233	Surgical	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem		
							F26	No Health Consequences or Impact	A24	No Apparent Device Problem		
					CT2406	Implantable	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem		
							F26	No Health Consequences or Impact	A24	No Apparent Device Problem		
					CT2407	Active implantable	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem		
							F26	No Health Consequences or Impact	A24	No Apparent Device Problem		
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem		
							F26	No Health Consequences or Impact	A24	No Apparent Device Problem		
					CT2727	Active implant devices/ancillaries	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem		
							F26	No Health Consequences or Impact	A24	No Apparent Device Problem		
					CT2750	Implantable cardiac electrical stimulators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem		
							F26	No Health Consequences or Impact	A24	No Apparent Device Problem		
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem		
							F26	No Health Consequences or Impact	A24	No Apparent Device Problem		
					CT2985	Active	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem		
							F26	No Health Consequences or Impact	A24	No Apparent Device Problem		
					CT3001	Implantable metal	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem		
							F26	No Health Consequences or Impact	A24	No Apparent Device Problem		
					CT3002	Implantable synthetic polymer	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem		
							F26	No Health Consequences or Impact	A24	No Apparent Device Problem		
					CT3006	Implantable inorganic material	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem		
							F26	No Health Consequences or Impact	A24	No Apparent Device Problem		
					CT314	Electrically powered	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem		
							F26	No Health Consequences or Impact	A24	No Apparent Device Problem		
					CT321	Long-term surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem		
							F26	No Health Consequences or Impact	A24	No Apparent Device Problem		
					CT449	Generators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem		
							F26	No Health Consequences or Impact	A24	No Apparent Device Problem		
					CT981	Single use	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem		
							F26	No Health Consequences or Impact	A24	No Apparent Device Problem		
					CT983	Surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem		
							F26	No Health Consequences or Impact	A24	No Apparent Device Problem		
13726	2022	All other reportable incidents	35852	Single-chamber implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	26/10/2022	26/10/2022
							F1901	Additional Surgery	A070504	Premature Discharge of Battery	26/10/2022	26/10/2022

13727	2022	All other reportable incidents	47265	Dual-chamber implantable pacemaker, rate-responsive	CT141	Electrodes/Leads and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	29/09/2022	29/09/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	29/09/2022	29/09/2022
					CT145	Cardiovascular devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	29/09/2022	29/09/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	29/09/2022	29/09/2022
					CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	29/09/2022	29/09/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	29/09/2022	29/09/2022
					CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	29/09/2022	29/09/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	29/09/2022	29/09/2022
					CT233	Surgical	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	29/09/2022	29/09/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	29/09/2022	29/09/2022
					CT2406	Implantable	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	29/09/2022	29/09/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	29/09/2022	29/09/2022
					CT2407	Active implantable	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	29/09/2022	29/09/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	29/09/2022	29/09/2022
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	29/09/2022	29/09/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	29/09/2022	29/09/2022
					CT2727	Active implant devices/ancillaries	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	29/09/2022	29/09/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	29/09/2022	29/09/2022
					CT2750	Implantable cardiac electrical stimulators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	29/09/2022	29/09/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	29/09/2022	29/09/2022
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	29/09/2022	29/09/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	29/09/2022	29/09/2022
					CT2985	Active	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	29/09/2022	29/09/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	29/09/2022	29/09/2022
					CT3001	Implantable metal	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	29/09/2022	29/09/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	29/09/2022	29/09/2022
					CT3002	Implantable synthetic polymer	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	29/09/2022	29/09/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	29/09/2022	29/09/2022
					CT3006	Implantable inorganic material	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	29/09/2022	29/09/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	29/09/2022	29/09/2022
					CT314	Electrically powered	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	29/09/2022	29/09/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	29/09/2022	29/09/2022
					CT321	Long-term surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	29/09/2022	29/09/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	29/09/2022	29/09/2022
					CT449	Generators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	29/09/2022	29/09/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	29/09/2022	29/09/2022
					CT981	Single use	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	29/09/2022	29/09/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	29/09/2022	29/09/2022
					CT983	Surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	29/09/2022	29/09/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	29/09/2022	29/09/2022
13728	2022	All other reportable incidents	47265	Dual-chamber implantable pacemaker, rate-responsive	CT141	Electrodes/Leads and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	27/09/2022	27/09/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	27/09/2022	27/09/2022
					CT145	Cardiovascular devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	27/09/2022	27/09/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	27/09/2022	27/09/2022
					CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	27/09/2022	27/09/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	27/09/2022	27/09/2022
					CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	27/09/2022	27/09/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	27/09/2022	27/09/2022
					CT233	Surgical	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	27/09/2022	27/09/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	27/09/2022	27/09/2022
					CT2406	Implantable	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	27/09/2022	27/09/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	27/09/2022	27/09/2022
					CT2407	Active implantable	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	27/09/2022	27/09/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	27/09/2022	27/09/2022
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	27/09/2022	27/09/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	27/09/2022	27/09/2022
					CT2727	Active implant devices/ancillaries	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	27/09/2022	27/09/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	27/09/2022	27/09/2022
					CT2750	Implantable cardiac electrical stimulators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	27/09/2022	27/09/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	27/09/2022	27/09/2022
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	27/09/2022	27/09/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	27/09/2022	27/09/2022
					CT2985	Active	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	27/09/2022	27/09/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	27/09/2022	27/09/2022
					CT3001	Implantable metal	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	27/09/2022	27/09/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	27/09/2022	27/09/2022
					CT3002	Implantable synthetic polymer	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	27/09/2022	27/09/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	27/09/2022	27/09/2022
					CT3006	Implantable inorganic material	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	27/09/2022	27/09/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	27/09/2022	27/09/2022
					CT314	Electrically powered	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	27/09/2022	27/09/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	27/09/2022	27/09/2022
					CT321	Long-term surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	27/09/2022	27/09/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	27/09/2022	27/09/2022
					CT449	Generators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	27/09/2022	27/09/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	27/09/2022	27/09/2022
					CT981	Single use	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	27/09/2022	27/09/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	27/09/2022	27/09/2022
					CT983	Surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	27/09/2022	27/09/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	27/09/2022	27/09/2022
13729	2022	All other reportable incidents	47263	Cardiac resynchronization therapy implantable pacemaker	CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		

13730	2022	All other reportable incidents	37265	Dual-chamber implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A070101	Failure to Capture	08/11/2022	08/11/2022
									A070908	Failure to Sense	08/11/2022	08/11/2022
									A072201	High Impedance	08/11/2022	08/11/2022
							F1901	Additional Surgery	A070101	Failure to Capture	08/11/2022	08/11/2022
									A070908	Failure to Sense	08/11/2022	08/11/2022
									A072201	High Impedance	08/11/2022	08/11/2022
13731	2022	All other reportable incidents	47265	Dual-chamber implantable pacemaker, rate-responsive	CT141	Electrodes/Leads and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	06/11/2022	06/11/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	06/11/2022	06/11/2022
					CT145	Cardiovascular devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	06/11/2022	06/11/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	06/11/2022	06/11/2022
					CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	06/11/2022	06/11/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	06/11/2022	06/11/2022
					CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	06/11/2022	06/11/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	06/11/2022	06/11/2022
					CT233	Surgical	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	06/11/2022	06/11/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	06/11/2022	06/11/2022
					CT2406	Implantable	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	06/11/2022	06/11/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	06/11/2022	06/11/2022
					CT2407	Active implantable	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	06/11/2022	06/11/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	06/11/2022	06/11/2022
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	06/11/2022	06/11/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	06/11/2022	06/11/2022
					CT2727	Active implant devices/ancillaries	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	06/11/2022	06/11/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	06/11/2022	06/11/2022
					CT2750	Implantable cardiac electrical stimulators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	06/11/2022	06/11/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	06/11/2022	06/11/2022
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	06/11/2022	06/11/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	06/11/2022	06/11/2022
					CT2985	Active	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	06/11/2022	06/11/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	06/11/2022	06/11/2022
					CT3001	Implantable metal	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	06/11/2022	06/11/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	06/11/2022	06/11/2022
					CT3002	Implantable synthetic polymer	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	06/11/2022	06/11/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	06/11/2022	06/11/2022
					CT3006	Implantable inorganic material	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	06/11/2022	06/11/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	06/11/2022	06/11/2022
					CT314	Electrically powered	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	06/11/2022	06/11/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	06/11/2022	06/11/2022
					CT321	Long-term surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	06/11/2022	06/11/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	06/11/2022	06/11/2022
					CT449	Generators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	06/11/2022	06/11/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	06/11/2022	06/11/2022
					CT981	Single use	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	06/11/2022	06/11/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	06/11/2022	06/11/2022
					CT983	Surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	06/11/2022	06/11/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	06/11/2022	06/11/2022
13732	2022	All other reportable incidents	47265	Dual-chamber implantable pacemaker, rate-responsive	CT141	Electrodes/Leads and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	06/11/2022	06/11/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	06/11/2022	06/11/2022
					CT145	Cardiovascular devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	06/11/2022	06/11/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	06/11/2022	06/11/2022
					CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	06/11/2022	06/11/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	06/11/2022	06/11/2022
					CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	06/11/2022	06/11/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	06/11/2022	06/11/2022
					CT233	Surgical	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	06/11/2022	06/11/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	06/11/2022	06/11/2022
					CT2406	Implantable	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	06/11/2022	06/11/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	06/11/2022	06/11/2022
					CT2407	Active implantable	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	06/11/2022	06/11/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	06/11/2022	06/11/2022
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	06/11/2022	06/11/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	06/11/2022	06/11/2022
					CT2727	Active implant devices/ancillaries	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	06/11/2022	06/11/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	06/11/2022	06/11/2022
					CT2750	Implantable cardiac electrical stimulators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	06/11/2022	06/11/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	06/11/2022	06/11/2022
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	06/11/2022	06/11/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	06/11/2022	06/11/2022
					CT2985	Active	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	06/11/2022	06/11/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	06/11/2022	06/11/2022
					CT3001	Implantable metal	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	06/11/2022	06/11/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	06/11/2022	06/11/2022
					CT3002	Implantable synthetic polymer	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	06/11/2022	06/11/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	06/11/2022	06/11/2022
					CT3006	Implantable inorganic material	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	06/11/2022	06/11/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	06/11/2022	06/11/2022
					CT314	Electrically powered	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	06/11/2022	06/11/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	06/11/2022	06/11/2022
					CT321	Long-term surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	06/11/2022	06/11/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	06/11/2022	06/11/2022
					CT449	Generators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	06/11/2022	06/11/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	06/11/2022	06/11/2022
					CT981	Single use	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	06/11/2022	06/11/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	06/11/2022	06/11/2022
					CT983	Surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	06/11/2022	06/11/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	06/11/2022	06/11/2022
13733	2022	All other reportable incidents	47265	Dual-chamber implantable pacemaker, rate-responsive	CT141	Electrodes/Leads and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	09/11/2022	09/11/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	09/11/2022	09/11/2022

					CT145	Cardiovascular devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	09/11/2022	09/11/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	09/11/2022	09/11/2022
					CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	09/11/2022	09/11/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	09/11/2022	09/11/2022
					CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	09/11/2022	09/11/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	09/11/2022	09/11/2022
					CT233	Surgical	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	09/11/2022	09/11/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	09/11/2022	09/11/2022
					CT2406	Implantable	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	09/11/2022	09/11/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	09/11/2022	09/11/2022
					CT2407	Active implantable	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	09/11/2022	09/11/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	09/11/2022	09/11/2022
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	09/11/2022	09/11/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	09/11/2022	09/11/2022
					CT2727	Active implant devices/ancillaries	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	09/11/2022	09/11/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	09/11/2022	09/11/2022
					CT2750	Implantable cardiac electrical stimulators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	09/11/2022	09/11/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	09/11/2022	09/11/2022
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	09/11/2022	09/11/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	09/11/2022	09/11/2022
					CT2985	Active	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	09/11/2022	09/11/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	09/11/2022	09/11/2022
					CT3001	Implantable metal	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	09/11/2022	09/11/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	09/11/2022	09/11/2022
					CT3002	Implantable synthetic polymer	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	09/11/2022	09/11/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	09/11/2022	09/11/2022
					CT3006	Implantable inorganic material	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	09/11/2022	09/11/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	09/11/2022	09/11/2022
					CT314	Electrically powered	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	09/11/2022	09/11/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	09/11/2022	09/11/2022
					CT321	Long-term surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	09/11/2022	09/11/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	09/11/2022	09/11/2022
					CT449	Generators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	09/11/2022	09/11/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	09/11/2022	09/11/2022
					CT981	Single use	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	09/11/2022	09/11/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	09/11/2022	09/11/2022
					CT983	Surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	09/11/2022	09/11/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	09/11/2022	09/11/2022
13734	2022	All other reportable incidents	47265	Dual-chamber implantable pacemaker, rate-responsive	CT141	Electrodes/Leads and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	09/11/2022	09/11/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	09/11/2022	09/11/2022
					CT145	Cardiovascular devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	09/11/2022	09/11/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	09/11/2022	09/11/2022
					CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	09/11/2022	09/11/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	09/11/2022	09/11/2022
					CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	09/11/2022	09/11/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	09/11/2022	09/11/2022
					CT233	Surgical	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	09/11/2022	09/11/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	09/11/2022	09/11/2022
					CT2406	Implantable	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	09/11/2022	09/11/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	09/11/2022	09/11/2022
					CT2407	Active implantable	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	09/11/2022	09/11/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	09/11/2022	09/11/2022
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	09/11/2022	09/11/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	09/11/2022	09/11/2022
					CT2727	Active implant devices/ancillaries	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	09/11/2022	09/11/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	09/11/2022	09/11/2022
					CT2750	Implantable cardiac electrical stimulators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	09/11/2022	09/11/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	09/11/2022	09/11/2022
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	09/11/2022	09/11/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	09/11/2022	09/11/2022
					CT2985	Active	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	09/11/2022	09/11/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	09/11/2022	09/11/2022
					CT3001	Implantable metal	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	09/11/2022	09/11/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	09/11/2022	09/11/2022
					CT3002	Implantable synthetic polymer	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	09/11/2022	09/11/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	09/11/2022	09/11/2022
					CT3006	Implantable inorganic material	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	09/11/2022	09/11/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	09/11/2022	09/11/2022
					CT314	Electrically powered	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	09/11/2022	09/11/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	09/11/2022	09/11/2022
					CT321	Long-term surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	09/11/2022	09/11/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	09/11/2022	09/11/2022
					CT449	Generators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	09/11/2022	09/11/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	09/11/2022	09/11/2022
					CT981	Single use	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	09/11/2022	09/11/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	09/11/2022	09/11/2022
					CT983	Surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	09/11/2022	09/11/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	09/11/2022	09/11/2022
13735	2022	All other reportable incidents	47265	Dual-chamber implantable pacemaker, rate-responsive	CT141	Electrodes/Leads and associated devices	E2340	Wound Dehiscence	A24	No Apparent Device Problem	08/11/2022	08/11/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	08/11/2022	08/11/2022
					CT145	Cardiovascular devices	E2340	Wound Dehiscence	A24	No Apparent Device Problem	08/11/2022	08/11/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	08/11/2022	08/11/2022
					CT1684	Cardiac pacemakers and associated devices	E2340	Wound Dehiscence	A24	No Apparent Device Problem	08/11/2022	08/11/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	08/11/2022	08/11/2022
					CT1685	Cardiac pacemakers	E2340	Wound Dehiscence	A24	No Apparent Device Problem	08/11/2022	08/11/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	08/11/2022	08/11/2022
					CT233	Surgical	E2340	Wound Dehiscence	A24	No Apparent Device Problem	08/11/2022	08/11/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	08/11/2022	08/11/2022

					CT2406	Implantable	E2340	Wound Dehiscence	A24	No Apparent Device Problem	08/11/2022	08/11/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	08/11/2022	08/11/2022
					CT2407	Active implantable	E2340	Wound Dehiscence	A24	No Apparent Device Problem	08/11/2022	08/11/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	08/11/2022	08/11/2022
					CT257	Implantable cardiac pacemakers	E2340	Wound Dehiscence	A24	No Apparent Device Problem	08/11/2022	08/11/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	08/11/2022	08/11/2022
					CT2727	Active implant devices/ancillaries	E2340	Wound Dehiscence	A24	No Apparent Device Problem	08/11/2022	08/11/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	08/11/2022	08/11/2022
					CT2750	Implantable cardiac electrical stimulators and associated devices	E2340	Wound Dehiscence	A24	No Apparent Device Problem	08/11/2022	08/11/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	08/11/2022	08/11/2022
					CT2866	Implantable cardiac electrical stimulators	E2340	Wound Dehiscence	A24	No Apparent Device Problem	08/11/2022	08/11/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	08/11/2022	08/11/2022
					CT2985	Active	E2340	Wound Dehiscence	A24	No Apparent Device Problem	08/11/2022	08/11/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	08/11/2022	08/11/2022
					CT3001	Implantable metal	E2340	Wound Dehiscence	A24	No Apparent Device Problem	08/11/2022	08/11/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	08/11/2022	08/11/2022
					CT3002	Implantable synthetic polymer	E2340	Wound Dehiscence	A24	No Apparent Device Problem	08/11/2022	08/11/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	08/11/2022	08/11/2022
					CT3006	Implantable inorganic material	E2340	Wound Dehiscence	A24	No Apparent Device Problem	08/11/2022	08/11/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	08/11/2022	08/11/2022
					CT314	Electrically powered	E2340	Wound Dehiscence	A24	No Apparent Device Problem	08/11/2022	08/11/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	08/11/2022	08/11/2022
					CT321	Long-term surgical invasive	E2340	Wound Dehiscence	A24	No Apparent Device Problem	08/11/2022	08/11/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	08/11/2022	08/11/2022
					CT449	Generators and associated devices	E2340	Wound Dehiscence	A24	No Apparent Device Problem	08/11/2022	08/11/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	08/11/2022	08/11/2022
					CT981	Single use	E2340	Wound Dehiscence	A24	No Apparent Device Problem	08/11/2022	08/11/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	08/11/2022	08/11/2022
					CT983	Surgical invasive	E2340	Wound Dehiscence	A24	No Apparent Device Problem	08/11/2022	08/11/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	08/11/2022	08/11/2022
13736	2022	All other reportable incidents	47265	Dual-chamber implantable pacemaker, rate-responsive	CT141	Electrodes/Leads and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A1204	Failure to Disconnect	01/11/2022	01/11/2022
							F1901	Additional Surgery	A1204	Failure to Disconnect	01/11/2022	01/11/2022
					CT145	Cardiovascular devices	E2403	No Clinical Signs, Symptoms or Conditions	A1204	Failure to Disconnect	01/11/2022	01/11/2022
							F1901	Additional Surgery	A1204	Failure to Disconnect	01/11/2022	01/11/2022
					CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A1204	Failure to Disconnect	01/11/2022	01/11/2022
							F1901	Additional Surgery	A1204	Failure to Disconnect	01/11/2022	01/11/2022
					CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A1204	Failure to Disconnect	01/11/2022	01/11/2022
							F1901	Additional Surgery	A1204	Failure to Disconnect	01/11/2022	01/11/2022
					CT233	Surgical	E2403	No Clinical Signs, Symptoms or Conditions	A1204	Failure to Disconnect	01/11/2022	01/11/2022
							F1901	Additional Surgery	A1204	Failure to Disconnect	01/11/2022	01/11/2022
					CT2406	Implantable	E2403	No Clinical Signs, Symptoms or Conditions	A1204	Failure to Disconnect	01/11/2022	01/11/2022
							F1901	Additional Surgery	A1204	Failure to Disconnect	01/11/2022	01/11/2022
					CT2407	Active implantable	E2403	No Clinical Signs, Symptoms or Conditions	A1204	Failure to Disconnect	01/11/2022	01/11/2022
							F1901	Additional Surgery	A1204	Failure to Disconnect	01/11/2022	01/11/2022
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A1204	Failure to Disconnect	01/11/2022	01/11/2022
							F1901	Additional Surgery	A1204	Failure to Disconnect	01/11/2022	01/11/2022
					CT2727	Active implant devices/ancillaries	E2403	No Clinical Signs, Symptoms or Conditions	A1204	Failure to Disconnect	01/11/2022	01/11/2022
							F1901	Additional Surgery	A1204	Failure to Disconnect	01/11/2022	01/11/2022
					CT2750	Implantable cardiac electrical stimulators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A1204	Failure to Disconnect	01/11/2022	01/11/2022
							F1901	Additional Surgery	A1204	Failure to Disconnect	01/11/2022	01/11/2022
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A1204	Failure to Disconnect	01/11/2022	01/11/2022
							F1901	Additional Surgery	A1204	Failure to Disconnect	01/11/2022	01/11/2022
					CT2985	Active	E2403	No Clinical Signs, Symptoms or Conditions	A1204	Failure to Disconnect	01/11/2022	01/11/2022
							F1901	Additional Surgery	A1204	Failure to Disconnect	01/11/2022	01/11/2022
					CT3001	Implantable metal	E2403	No Clinical Signs, Symptoms or Conditions	A1204	Failure to Disconnect	01/11/2022	01/11/2022
							F1901	Additional Surgery	A1204	Failure to Disconnect	01/11/2022	01/11/2022
					CT3002	Implantable synthetic polymer	E2403	No Clinical Signs, Symptoms or Conditions	A1204	Failure to Disconnect	01/11/2022	01/11/2022
							F1901	Additional Surgery	A1204	Failure to Disconnect	01/11/2022	01/11/2022
					CT3006	Implantable inorganic material	E2403	No Clinical Signs, Symptoms or Conditions	A1204	Failure to Disconnect	01/11/2022	01/11/2022
							F1901	Additional Surgery	A1204	Failure to Disconnect	01/11/2022	01/11/2022
					CT314	Electrically powered	E2403	No Clinical Signs, Symptoms or Conditions	A1204	Failure to Disconnect	01/11/2022	01/11/2022
							F1901	Additional Surgery	A1204	Failure to Disconnect	01/11/2022	01/11/2022
					CT321	Long-term surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A1204	Failure to Disconnect	01/11/2022	01/11/2022
							F1901	Additional Surgery	A1204	Failure to Disconnect	01/11/2022	01/11/2022
					CT449	Generators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A1204	Failure to Disconnect	01/11/2022	01/11/2022
							F1901	Additional Surgery	A1204	Failure to Disconnect	01/11/2022	01/11/2022
					CT981	Single use	E2403	No Clinical Signs, Symptoms or Conditions	A1204	Failure to Disconnect	01/11/2022	01/11/2022
							F1901	Additional Surgery	A1204	Failure to Disconnect	01/11/2022	01/11/2022
					CT983	Surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A1204	Failure to Disconnect	01/11/2022	01/11/2022
							F1901	Additional Surgery	A1204	Failure to Disconnect	01/11/2022	01/11/2022
13737	2022	All other reportable incidents	47265	Dual-chamber implantable pacemaker, rate-responsive	CT141	Electrodes/Leads and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	09/11/2022	09/11/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	09/11/2022	09/11/2022
					CT145	Cardiovascular devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	09/11/2022	09/11/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	09/11/2022	09/11/2022
					CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	09/11/2022	09/11/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	09/11/2022	09/11/2022
					CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	09/11/2022	09/11/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	09/11/2022	09/11/2022
					CT233	Surgical	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	09/11/2022	09/11/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	09/11/2022	09/11/2022
					CT2406	Implantable	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	09/11/2022	09/11/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	09/11/2022	09/11/2022
					CT2407	Active implantable	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	09/11/2022	09/11/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	09/11/2022	09/11/2022
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	09/11/2022	09/11/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	09/11/2022	09/11/2022
					CT2727	Active implant devices/ancillaries	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	09/11/2022	09/11/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	09/11/2022	09/11/2022

					CT2750	Implantable cardiac electrical stimulators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	09/11/2022	09/11/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	09/11/2022	09/11/2022
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	09/11/2022	09/11/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	09/11/2022	09/11/2022
					CT2985	Active	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	09/11/2022	09/11/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	09/11/2022	09/11/2022
					CT3001	Implantable metal	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	09/11/2022	09/11/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	09/11/2022	09/11/2022
					CT3002	Implantable synthetic polymer	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	09/11/2022	09/11/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	09/11/2022	09/11/2022
					CT3006	Implantable inorganic material	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	09/11/2022	09/11/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	09/11/2022	09/11/2022
					CT314	Electrically powered	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	09/11/2022	09/11/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	09/11/2022	09/11/2022
					CT321	Long-term surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	09/11/2022	09/11/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	09/11/2022	09/11/2022
					CT449	Generators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	09/11/2022	09/11/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	09/11/2022	09/11/2022
					CT981	Single use	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	09/11/2022	09/11/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	09/11/2022	09/11/2022
					CT983	Surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	09/11/2022	09/11/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	09/11/2022	09/11/2022
13738	2022	All other reportable incidents	47265	Dual-chamber implantable pacemaker, rate-responsive	CT141	Electrodes/Leads and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem		
							F1901	Additional Surgery	A24	No Apparent Device Problem		
					CT145	Cardiovascular devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem		
							F1901	Additional Surgery	A24	No Apparent Device Problem		
					CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem		
							F1901	Additional Surgery	A24	No Apparent Device Problem		
					CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem		
							F1901	Additional Surgery	A24	No Apparent Device Problem		
					CT233	Surgical	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem		
							F1901	Additional Surgery	A24	No Apparent Device Problem		
					CT2406	Implantable	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem		
							F1901	Additional Surgery	A24	No Apparent Device Problem		
					CT2407	Active implantable	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem		
							F1901	Additional Surgery	A24	No Apparent Device Problem		
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem		
							F1901	Additional Surgery	A24	No Apparent Device Problem		
					CT2727	Active implant devices/ancillaries	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem		
							F1901	Additional Surgery	A24	No Apparent Device Problem		
					CT2750	Implantable cardiac electrical stimulators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem		
							F1901	Additional Surgery	A24	No Apparent Device Problem		
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem		
							F1901	Additional Surgery	A24	No Apparent Device Problem		
					CT2985	Active	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem		
							F1901	Additional Surgery	A24	No Apparent Device Problem		
					CT3001	Implantable metal	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem		
							F1901	Additional Surgery	A24	No Apparent Device Problem		
					CT3002	Implantable synthetic polymer	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem		
							F1901	Additional Surgery	A24	No Apparent Device Problem		
					CT3006	Implantable inorganic material	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem		
							F1901	Additional Surgery	A24	No Apparent Device Problem		
					CT314	Electrically powered	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem		
							F1901	Additional Surgery	A24	No Apparent Device Problem		
					CT321	Long-term surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem		
							F1901	Additional Surgery	A24	No Apparent Device Problem		
					CT449	Generators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem		
							F1901	Additional Surgery	A24	No Apparent Device Problem		
					CT981	Single use	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem		
							F1901	Additional Surgery	A24	No Apparent Device Problem		
					CT983	Surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem		
							F1901	Additional Surgery	A24	No Apparent Device Problem		
13739	2022	All other reportable incidents	47265	Dual-chamber implantable pacemaker, rate-responsive	CT141	Electrodes/Leads and associated devices	E0601	Arrhythmia	A071204	No Pacing	02/11/2022	02/11/2022
							F1908	Prolonged Surgery	A071204	No Pacing	02/11/2022	02/11/2022
					CT145	Cardiovascular devices	E0601	Arrhythmia	A071204	No Pacing	02/11/2022	02/11/2022
							F1908	Prolonged Surgery	A071204	No Pacing	02/11/2022	02/11/2022
					CT1684	Cardiac pacemakers and associated devices	E0601	Arrhythmia	A071204	No Pacing	02/11/2022	02/11/2022
							F1908	Prolonged Surgery	A071204	No Pacing	02/11/2022	02/11/2022
					CT1685	Cardiac pacemakers	E0601	Arrhythmia	A071204	No Pacing	02/11/2022	02/11/2022
							F1908	Prolonged Surgery	A071204	No Pacing	02/11/2022	02/11/2022
					CT233	Surgical	E0601	Arrhythmia	A071204	No Pacing	02/11/2022	02/11/2022
							F1908	Prolonged Surgery	A071204	No Pacing	02/11/2022	02/11/2022
					CT2406	Implantable	E0601	Arrhythmia	A071204	No Pacing	02/11/2022	02/11/2022
							F1908	Prolonged Surgery	A071204	No Pacing	02/11/2022	02/11/2022
					CT2407	Active implantable	E0601	Arrhythmia	A071204	No Pacing	02/11/2022	02/11/2022
							F1908	Prolonged Surgery	A071204	No Pacing	02/11/2022	02/11/2022
					CT257	Implantable cardiac pacemakers	E0601	Arrhythmia	A071204	No Pacing	02/11/2022	02/11/2022
							F1908	Prolonged Surgery	A071204	No Pacing	02/11/2022	02/11/2022
					CT2727	Active implant devices/ancillaries	E0601	Arrhythmia	A071204	No Pacing	02/11/2022	02/11/2022
							F1908	Prolonged Surgery	A071204	No Pacing	02/11/2022	02/11/2022
					CT2750	Implantable cardiac electrical stimulators and associated devices	E0601	Arrhythmia	A071204	No Pacing	02/11/2022	02/11/2022
							F1908	Prolonged Surgery	A071204	No Pacing	02/11/2022	02/11/2022
					CT2866	Implantable cardiac electrical stimulators	E0601	Arrhythmia	A071204	No Pacing	02/11/2022	02/11/2022
							F1908	Prolonged Surgery	A071204	No Pacing	02/11/2022	02/11/2022
					CT2985	Active	E0601	Arrhythmia	A071204	No Pacing	02/11/2022	02/11/2022
							F1908	Prolonged Surgery	A071204	No Pacing	02/11/2022	02/11/2022
					CT3001	Implantable metal	E0601	Arrhythmia	A071204	No Pacing	02/11/2022	02/11/2022
							F1908	Prolonged Surgery	A071204	No Pacing	02/11/2022	02/11/2022

					CT3002	Implantable synthetic polymer	E0601	Arrhythmia	A071204	No Pacing	02/11/2022	02/11/2022
							F1908	Prolonged Surgery	A071204	No Pacing	02/11/2022	02/11/2022
					CT3006	Implantable inorganic material	E0601	Arrhythmia	A071204	No Pacing	02/11/2022	02/11/2022
							F1908	Prolonged Surgery	A071204	No Pacing	02/11/2022	02/11/2022
					CT314	Electrically powered	E0601	Arrhythmia	A071204	No Pacing	02/11/2022	02/11/2022
							F1908	Prolonged Surgery	A071204	No Pacing	02/11/2022	02/11/2022
					CT321	Long-term surgical invasive	E0601	Arrhythmia	A071204	No Pacing	02/11/2022	02/11/2022
							F1908	Prolonged Surgery	A071204	No Pacing	02/11/2022	02/11/2022
					CT449	Generators and associated devices	E0601	Arrhythmia	A071204	No Pacing	02/11/2022	02/11/2022
							F1908	Prolonged Surgery	A071204	No Pacing	02/11/2022	02/11/2022
					CT981	Single use	E0601	Arrhythmia	A071204	No Pacing	02/11/2022	02/11/2022
							F1908	Prolonged Surgery	A071204	No Pacing	02/11/2022	02/11/2022
					CT983	Surgical invasive	E0601	Arrhythmia	A071204	No Pacing	02/11/2022	02/11/2022
							F1908	Prolonged Surgery	A071204	No Pacing	02/11/2022	02/11/2022
13740	2022	All other reportable incidents	47265	Dual-chamber implantable pacemaker, rate-responsive	CT141	Electrodes/Leads and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem		
							F26	No Health Consequences or Impact	A24	No Apparent Device Problem		
					CT145	Cardiovascular devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem		
							F26	No Health Consequences or Impact	A24	No Apparent Device Problem		
					CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem		
							F26	No Health Consequences or Impact	A24	No Apparent Device Problem		
					CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem		
							F26	No Health Consequences or Impact	A24	No Apparent Device Problem		
					CT233	Surgical	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem		
							F26	No Health Consequences or Impact	A24	No Apparent Device Problem		
					CT2406	Implantable	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem		
							F26	No Health Consequences or Impact	A24	No Apparent Device Problem		
					CT2407	Active implantable	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem		
							F26	No Health Consequences or Impact	A24	No Apparent Device Problem		
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem		
							F26	No Health Consequences or Impact	A24	No Apparent Device Problem		
					CT2727	Active implant devices/ancillaries	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem		
							F26	No Health Consequences or Impact	A24	No Apparent Device Problem		
					CT2750	Implantable cardiac electrical stimulators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem		
							F26	No Health Consequences or Impact	A24	No Apparent Device Problem		
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem		
							F26	No Health Consequences or Impact	A24	No Apparent Device Problem		
					CT2985	Active	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem		
							F26	No Health Consequences or Impact	A24	No Apparent Device Problem		
					CT3001	Implantable metal	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem		
							F26	No Health Consequences or Impact	A24	No Apparent Device Problem		
					CT3002	Implantable synthetic polymer	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem		
							F26	No Health Consequences or Impact	A24	No Apparent Device Problem		
					CT3006	Implantable inorganic material	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem		
							F26	No Health Consequences or Impact	A24	No Apparent Device Problem		
					CT314	Electrically powered	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem		
							F26	No Health Consequences or Impact	A24	No Apparent Device Problem		
					CT321	Long-term surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem		
							F26	No Health Consequences or Impact	A24	No Apparent Device Problem		
					CT449	Generators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem		
							F26	No Health Consequences or Impact	A24	No Apparent Device Problem		
					CT981	Single use	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem		
							F26	No Health Consequences or Impact	A24	No Apparent Device Problem		
					CT983	Surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem		
							F26	No Health Consequences or Impact	A24	No Apparent Device Problem		
13741	2022	All other reportable incidents	47265	Dual-chamber implantable pacemaker, rate-responsive	CT141	Electrodes/Leads and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	01/10/2022	01/10/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	01/10/2022	01/10/2022
					CT145	Cardiovascular devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	01/10/2022	01/10/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	01/10/2022	01/10/2022
					CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	01/10/2022	01/10/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	01/10/2022	01/10/2022
					CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	01/10/2022	01/10/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	01/10/2022	01/10/2022
					CT233	Surgical	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	01/10/2022	01/10/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	01/10/2022	01/10/2022
					CT2406	Implantable	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	01/10/2022	01/10/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	01/10/2022	01/10/2022
					CT2407	Active implantable	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	01/10/2022	01/10/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	01/10/2022	01/10/2022
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	01/10/2022	01/10/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	01/10/2022	01/10/2022
					CT2727	Active implant devices/ancillaries	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	01/10/2022	01/10/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	01/10/2022	01/10/2022
					CT2750	Implantable cardiac electrical stimulators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	01/10/2022	01/10/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	01/10/2022	01/10/2022
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	01/10/2022	01/10/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	01/10/2022	01/10/2022
					CT2985	Active	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	01/10/2022	01/10/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	01/10/2022	01/10/2022
					CT3001	Implantable metal	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	01/10/2022	01/10/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	01/10/2022	01/10/2022
					CT3002	Implantable synthetic polymer	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	01/10/2022	01/10/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	01/10/2022	01/10/2022
					CT3006	Implantable inorganic material	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	01/10/2022	01/10/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	01/10/2022	01/10/2022
					CT314	Electrically powered	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	01/10/2022	01/10/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	01/10/2022	01/10/2022
					CT321	Long-term surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	01/10/2022	01/10/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	01/10/2022	01/10/2022

					CT449	Generators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	01/10/2022	01/10/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	01/10/2022	01/10/2022
					CT981	Single use	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	01/10/2022	01/10/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	01/10/2022	01/10/2022
					CT983	Surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	01/10/2022	01/10/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	01/10/2022	01/10/2022
13742	2022	All other reportable incidents	47265	Dual-chamber implantable pacemaker, rate-responsive	CT141	Electrodes/Leads and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	18/11/2022	18/11/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	18/11/2022	18/11/2022
					CT145	Cardiovascular devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	18/11/2022	18/11/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	18/11/2022	18/11/2022
					CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	18/11/2022	18/11/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	18/11/2022	18/11/2022
					CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	18/11/2022	18/11/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	18/11/2022	18/11/2022
					CT233	Surgical	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	18/11/2022	18/11/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	18/11/2022	18/11/2022
					CT2406	Implantable	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	18/11/2022	18/11/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	18/11/2022	18/11/2022
					CT2407	Active implantable	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	18/11/2022	18/11/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	18/11/2022	18/11/2022
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	18/11/2022	18/11/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	18/11/2022	18/11/2022
					CT2727	Active implant devices/ancillaries	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	18/11/2022	18/11/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	18/11/2022	18/11/2022
					CT2750	Implantable cardiac electrical stimulators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	18/11/2022	18/11/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	18/11/2022	18/11/2022
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	18/11/2022	18/11/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	18/11/2022	18/11/2022
					CT2985	Active	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	18/11/2022	18/11/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	18/11/2022	18/11/2022
					CT3001	Implantable metal	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	18/11/2022	18/11/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	18/11/2022	18/11/2022
					CT3002	Implantable synthetic polymer	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	18/11/2022	18/11/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	18/11/2022	18/11/2022
					CT3006	Implantable inorganic material	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	18/11/2022	18/11/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	18/11/2022	18/11/2022
					CT314	Electrically powered	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	18/11/2022	18/11/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	18/11/2022	18/11/2022
					CT321	Long-term surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	18/11/2022	18/11/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	18/11/2022	18/11/2022
					CT449	Generators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	18/11/2022	18/11/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	18/11/2022	18/11/2022
					CT981	Single use	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	18/11/2022	18/11/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	18/11/2022	18/11/2022
					CT983	Surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	18/11/2022	18/11/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	18/11/2022	18/11/2022
13743	2022	All other reportable incidents	47265	Dual-chamber implantable pacemaker, rate-responsive	CT141	Electrodes/Leads and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	09/11/2022	09/11/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	09/11/2022	09/11/2022
					CT145	Cardiovascular devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	09/11/2022	09/11/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	09/11/2022	09/11/2022
					CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	09/11/2022	09/11/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	09/11/2022	09/11/2022
					CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	09/11/2022	09/11/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	09/11/2022	09/11/2022
					CT233	Surgical	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	09/11/2022	09/11/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	09/11/2022	09/11/2022
					CT2406	Implantable	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	09/11/2022	09/11/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	09/11/2022	09/11/2022
					CT2407	Active implantable	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	09/11/2022	09/11/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	09/11/2022	09/11/2022
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	09/11/2022	09/11/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	09/11/2022	09/11/2022
					CT2727	Active implant devices/ancillaries	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	09/11/2022	09/11/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	09/11/2022	09/11/2022
					CT2750	Implantable cardiac electrical stimulators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	09/11/2022	09/11/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	09/11/2022	09/11/2022
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	09/11/2022	09/11/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	09/11/2022	09/11/2022
					CT2985	Active	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	09/11/2022	09/11/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	09/11/2022	09/11/2022
					CT3001	Implantable metal	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	09/11/2022	09/11/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	09/11/2022	09/11/2022
					CT3002	Implantable synthetic polymer	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	09/11/2022	09/11/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	09/11/2022	09/11/2022
					CT3006	Implantable inorganic material	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	09/11/2022	09/11/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	09/11/2022	09/11/2022
					CT314	Electrically powered	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	09/11/2022	09/11/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	09/11/2022	09/11/2022
					CT321	Long-term surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	09/11/2022	09/11/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	09/11/2022	09/11/2022
					CT449	Generators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	09/11/2022	09/11/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	09/11/2022	09/11/2022
					CT981	Single use	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	09/11/2022	09/11/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	09/11/2022	09/11/2022
					CT983	Surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	09/11/2022	09/11/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	09/11/2022	09/11/2022
13744	2022		47265	Dual-chamber implantable pacemaker, rate-responsive	CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	22/11/2022	22/11/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	22/11/2022	22/11/2022

					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	22/11/2022	22/11/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	22/11/2022	22/11/2022
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	22/11/2022	22/11/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	22/11/2022	22/11/2022
13745	2023		47265	Dual-chamber implantable pacemaker, rate-responsive	CT141	Electrodes/Leads and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	07/06/2023	07/06/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	07/06/2023	07/06/2023
					CT145	Cardiovascular devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	07/06/2023	07/06/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	07/06/2023	07/06/2023
					CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	07/06/2023	07/06/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	07/06/2023	07/06/2023
					CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	07/06/2023	07/06/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	07/06/2023	07/06/2023
					CT233	Surgical	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	07/06/2023	07/06/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	07/06/2023	07/06/2023
					CT2406	Implantable	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	07/06/2023	07/06/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	07/06/2023	07/06/2023
					CT2407	Active implantable	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	07/06/2023	07/06/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	07/06/2023	07/06/2023
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	07/06/2023	07/06/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	07/06/2023	07/06/2023
					CT2727	Active implant devices/ancillaries	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	07/06/2023	07/06/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	07/06/2023	07/06/2023
					CT2750	Implantable cardiac electrical stimulators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	07/06/2023	07/06/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	07/06/2023	07/06/2023
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	07/06/2023	07/06/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	07/06/2023	07/06/2023
					CT2985	Active	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	07/06/2023	07/06/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	07/06/2023	07/06/2023
					CT3001	Implantable metal	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	07/06/2023	07/06/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	07/06/2023	07/06/2023
					CT3002	Implantable synthetic polymer	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	07/06/2023	07/06/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	07/06/2023	07/06/2023
					CT3006	Implantable inorganic material	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	07/06/2023	07/06/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	07/06/2023	07/06/2023
					CT314	Electrically powered	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	07/06/2023	07/06/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	07/06/2023	07/06/2023
					CT321	Long-term surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	07/06/2023	07/06/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	07/06/2023	07/06/2023
					CT449	Generators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	07/06/2023	07/06/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	07/06/2023	07/06/2023
					CT981	Single use	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	07/06/2023	07/06/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	07/06/2023	07/06/2023
					CT983	Surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	07/06/2023	07/06/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	07/06/2023	07/06/2023
13746	2023		47265	Dual-chamber implantable pacemaker, rate-responsive	CT141	Electrodes/Leads and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	29/06/2023	29/06/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	29/06/2023	29/06/2023
					CT145	Cardiovascular devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	29/06/2023	29/06/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	29/06/2023	29/06/2023
					CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	29/06/2023	29/06/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	29/06/2023	29/06/2023
					CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	29/06/2023	29/06/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	29/06/2023	29/06/2023
					CT233	Surgical	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	29/06/2023	29/06/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	29/06/2023	29/06/2023
					CT2406	Implantable	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	29/06/2023	29/06/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	29/06/2023	29/06/2023
					CT2407	Active implantable	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	29/06/2023	29/06/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	29/06/2023	29/06/2023
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	29/06/2023	29/06/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	29/06/2023	29/06/2023
					CT2727	Active implant devices/ancillaries	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	29/06/2023	29/06/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	29/06/2023	29/06/2023
					CT2750	Implantable cardiac electrical stimulators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	29/06/2023	29/06/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	29/06/2023	29/06/2023
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	29/06/2023	29/06/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	29/06/2023	29/06/2023
					CT2985	Active	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	29/06/2023	29/06/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	29/06/2023	29/06/2023
					CT3001	Implantable metal	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	29/06/2023	29/06/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	29/06/2023	29/06/2023
					CT3002	Implantable synthetic polymer	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	29/06/2023	29/06/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	29/06/2023	29/06/2023
					CT3006	Implantable inorganic material	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	29/06/2023	29/06/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	29/06/2023	29/06/2023
					CT314	Electrically powered	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	29/06/2023	29/06/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	29/06/2023	29/06/2023
					CT321	Long-term surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	29/06/2023	29/06/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	29/06/2023	29/06/2023
					CT449	Generators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	29/06/2023	29/06/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	29/06/2023	29/06/2023
					CT981	Single use	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	29/06/2023	29/06/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	29/06/2023	29/06/2023
					CT983	Surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	29/06/2023	29/06/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	29/06/2023	29/06/2023
13747	2023		47265	Dual-chamber implantable pacemaker, rate-responsive	CT141	Electrodes/Leads and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	22/06/2023	22/06/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	22/06/2023	22/06/2023
					CT145	Cardiovascular devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	22/06/2023	22/06/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	22/06/2023	22/06/2023

					CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	22/06/2023	22/06/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	22/06/2023	22/06/2023
					CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	22/06/2023	22/06/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	22/06/2023	22/06/2023
					CT233	Surgical	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	22/06/2023	22/06/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	22/06/2023	22/06/2023
					CT2406	Implantable	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	22/06/2023	22/06/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	22/06/2023	22/06/2023
					CT2407	Active implantable	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	22/06/2023	22/06/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	22/06/2023	22/06/2023
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	22/06/2023	22/06/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	22/06/2023	22/06/2023
					CT2727	Active implant devices/ancillaries	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	22/06/2023	22/06/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	22/06/2023	22/06/2023
					CT2750	Implantable cardiac electrical stimulators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	22/06/2023	22/06/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	22/06/2023	22/06/2023
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	22/06/2023	22/06/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	22/06/2023	22/06/2023
					CT2985	Active	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	22/06/2023	22/06/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	22/06/2023	22/06/2023
					CT3001	Implantable metal	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	22/06/2023	22/06/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	22/06/2023	22/06/2023
					CT3002	Implantable synthetic polymer	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	22/06/2023	22/06/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	22/06/2023	22/06/2023
					CT3006	Implantable inorganic material	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	22/06/2023	22/06/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	22/06/2023	22/06/2023
					CT314	Electrically powered	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	22/06/2023	22/06/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	22/06/2023	22/06/2023
					CT321	Long-term surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	22/06/2023	22/06/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	22/06/2023	22/06/2023
					CT449	Generators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	22/06/2023	22/06/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	22/06/2023	22/06/2023
					CT981	Single use	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	22/06/2023	22/06/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	22/06/2023	22/06/2023
					CT983	Surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	22/06/2023	22/06/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	22/06/2023	22/06/2023
13748	2023		47265	Dual-chamber implantable pacemaker, rate-responsive	CT141	Electrodes/Leads and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	22/06/2023	22/06/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	22/06/2023	22/06/2023
					CT145	Cardiovascular devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	22/06/2023	22/06/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	22/06/2023	22/06/2023
					CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	22/06/2023	22/06/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	22/06/2023	22/06/2023
					CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	22/06/2023	22/06/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	22/06/2023	22/06/2023
					CT233	Surgical	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	22/06/2023	22/06/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	22/06/2023	22/06/2023
					CT2406	Implantable	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	22/06/2023	22/06/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	22/06/2023	22/06/2023
					CT2407	Active implantable	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	22/06/2023	22/06/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	22/06/2023	22/06/2023
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	22/06/2023	22/06/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	22/06/2023	22/06/2023
					CT2727	Active implant devices/ancillaries	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	22/06/2023	22/06/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	22/06/2023	22/06/2023
					CT2750	Implantable cardiac electrical stimulators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	22/06/2023	22/06/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	22/06/2023	22/06/2023
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	22/06/2023	22/06/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	22/06/2023	22/06/2023
					CT2985	Active	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	22/06/2023	22/06/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	22/06/2023	22/06/2023
					CT3001	Implantable metal	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	22/06/2023	22/06/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	22/06/2023	22/06/2023
					CT3002	Implantable synthetic polymer	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	22/06/2023	22/06/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	22/06/2023	22/06/2023
					CT3006	Implantable inorganic material	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	22/06/2023	22/06/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	22/06/2023	22/06/2023
					CT314	Electrically powered	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	22/06/2023	22/06/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	22/06/2023	22/06/2023
					CT321	Long-term surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	22/06/2023	22/06/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	22/06/2023	22/06/2023
					CT449	Generators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	22/06/2023	22/06/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	22/06/2023	22/06/2023
					CT981	Single use	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	22/06/2023	22/06/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	22/06/2023	22/06/2023
					CT983	Surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	22/06/2023	22/06/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	22/06/2023	22/06/2023
13749	2023		47265	Dual-chamber implantable pacemaker, rate-responsive	CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	21/02/2023	21/02/2023
							F1901	Additional Surgery	A070504	Premature Discharge of Battery	21/02/2023	21/02/2023
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	21/02/2023	21/02/2023
							F1901	Additional Surgery	A070504	Premature Discharge of Battery	21/02/2023	21/02/2023
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	21/02/2023	21/02/2023
							F1901	Additional Surgery	A070504	Premature Discharge of Battery	21/02/2023	21/02/2023
13750	2023		47265	Dual-chamber implantable pacemaker, rate-responsive	CT141	Electrodes/Leads and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem		
							F1901	Additional Surgery	A24	No Apparent Device Problem		
					CT145	Cardiovascular devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem		
							F1901	Additional Surgery	A24	No Apparent Device Problem		
					CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem		
							F1901	Additional Surgery	A24	No Apparent Device Problem		

					CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem		
							F1901	Additional Surgery	A24	No Apparent Device Problem		
					CT233	Surgical	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem		
							F1901	Additional Surgery	A24	No Apparent Device Problem		
					CT2406	Implantable	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem		
							F1901	Additional Surgery	A24	No Apparent Device Problem		
					CT2407	Active implantable	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem		
							F1901	Additional Surgery	A24	No Apparent Device Problem		
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem		
							F1901	Additional Surgery	A24	No Apparent Device Problem		
					CT2727	Active implant devices/ancillaries	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem		
							F1901	Additional Surgery	A24	No Apparent Device Problem		
					CT2750	Implantable cardiac electrical stimulators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem		
							F1901	Additional Surgery	A24	No Apparent Device Problem		
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem		
							F1901	Additional Surgery	A24	No Apparent Device Problem		
					CT2985	Active	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem		
							F1901	Additional Surgery	A24	No Apparent Device Problem		
					CT3001	Implantable metal	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem		
							F1901	Additional Surgery	A24	No Apparent Device Problem		
					CT3002	Implantable synthetic polymer	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem		
							F1901	Additional Surgery	A24	No Apparent Device Problem		
					CT3006	Implantable inorganic material	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem		
							F1901	Additional Surgery	A24	No Apparent Device Problem		
					CT314	Electrically powered	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem		
							F1901	Additional Surgery	A24	No Apparent Device Problem		
					CT321	Long-term surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem		
							F1901	Additional Surgery	A24	No Apparent Device Problem		
					CT449	Generators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem		
							F1901	Additional Surgery	A24	No Apparent Device Problem		
					CT981	Single use	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem		
							F1901	Additional Surgery	A24	No Apparent Device Problem		
					CT983	Surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem		
							F1901	Additional Surgery	A24	No Apparent Device Problem		
13751	2023	All other reportable incidents	47265	Dual-chamber implantable pacemaker, rate-responsive	CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	17/05/2023	17/05/2023
									A071205	Pacemaker Found in Back-Up Mode	17/05/2023	17/05/2023
							F1901	Additional Surgery	A070504	Premature Discharge of Battery	17/05/2023	17/05/2023
									A071205	Pacemaker Found in Back-Up Mode	17/05/2023	17/05/2023
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	17/05/2023	17/05/2023
									A071205	Pacemaker Found in Back-Up Mode	17/05/2023	17/05/2023
							F1901	Additional Surgery	A070504	Premature Discharge of Battery	17/05/2023	17/05/2023
									A071205	Pacemaker Found in Back-Up Mode	17/05/2023	17/05/2023
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	17/05/2023	17/05/2023
									A071205	Pacemaker Found in Back-Up Mode	17/05/2023	17/05/2023
							F1901	Additional Surgery	A070504	Premature Discharge of Battery	17/05/2023	17/05/2023
									A071205	Pacemaker Found in Back-Up Mode	17/05/2023	17/05/2023
13752	2023	All other reportable incidents	17882	Physiologic-monitoring defibrillation system	CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A0908	Incorrect, Inadequate or Imprecise Result or Readings		
							F26	No Health Consequences or Impact	A0908	Incorrect, Inadequate or Imprecise Result or Readings		
13753	2023	All other reportable incidents	17882	Physiologic-monitoring defibrillation system	CT1684	Cardiac pacemakers and associated devices	E060106	Heart Block	A071204	No Pacing		
							F26	No Health Consequences or Impact	A071204	No Pacing		
13754	2023	All other reportable incidents	35223	Endocardial/interventricular septal pacing lead	CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070909	Oversensing		
							F26	No Health Consequences or Impact	A070909	Oversensing		
13755	2023	All other reportable incidents	63188	Left-ventricular wireless pacing system sensor/transmitter	CT141	Electrodes/Leads and associated devices			A0708	Power Problem		
					CT145	Cardiovascular devices			A0708	Power Problem		
					CT1577	Wireless monitoring system transmitters			A0708	Power Problem		
					CT1684	Cardiac pacemakers and associated devices			A0708	Power Problem		
					CT2317	Transmitters/Receivers/Transceivers			A0708	Power Problem		
					CT233	Surgical			A0708	Power Problem		
					CT2406	Implantable			A0708	Power Problem		
					CT2407	Active implantable			A0708	Power Problem		
					CT255	Transducers and associated devices			A0708	Power Problem		
					CT2726	Physiological monitoring/measuring devices			A0708	Power Problem		
					CT2727	Active implant devices/ancillaries			A0708	Power Problem		
					CT2750	Implantable cardiac electrical stimulators and associated devices			A0708	Power Problem		
					CT2885	Physiological monitoring/measuring system transmitters/receivers/transceivers			A0708	Power Problem		
					CT2985	Active			A0708	Power Problem		
					CT3001	Implantable metal			A0708	Power Problem		
					CT3002	Implantable synthetic polymer			A0708	Power Problem		
					CT3006	Implantable inorganic material			A0708	Power Problem		
					CT314	Electrically powered			A0708	Power Problem		
					CT321	Long-term surgical invasive			A0708	Power Problem		
					CT384	Sensors and associated devices			A0708	Power Problem		
					CT449	Generators and associated devices			A0708	Power Problem		
					CT981	Single use			A0708	Power Problem		
					CT983	Surgical invasive			A0708	Power Problem		
					CT994	Sensors			A0708	Power Problem		
13756	2023		47270	Cardiac resynchronization therapy implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E2401	Insufficient Information	A24	No Apparent Device Problem		
							F02	Death	A24	No Apparent Device Problem		
13757	2023	All other reportable incidents	17882	Physiologic-monitoring defibrillation system	CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070908	Failure to Sense		
							F26	No Health Consequences or Impact	A070908	Failure to Sense		
13758	2023		47265	Dual-chamber implantable pacemaker, rate-responsive	CT141	Electrodes/Leads and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	26/02/2023	26/02/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	26/02/2023	26/02/2023
					CT145	Cardiovascular devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	26/02/2023	26/02/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	26/02/2023	26/02/2023
					CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	26/02/2023	26/02/2023

							F1901	Additional Surgery	A24	No Apparent Device Problem	26/02/2023	26/02/2023
					CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	26/02/2023	26/02/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	26/02/2023	26/02/2023
					CT233	Surgical	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	26/02/2023	26/02/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	26/02/2023	26/02/2023
					CT2406	Implantable	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	26/02/2023	26/02/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	26/02/2023	26/02/2023
					CT2407	Active implantable	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	26/02/2023	26/02/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	26/02/2023	26/02/2023
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	26/02/2023	26/02/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	26/02/2023	26/02/2023
					CT2727	Active implant devices/ancillaries	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	26/02/2023	26/02/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	26/02/2023	26/02/2023
					CT2750	Implantable cardiac electrical stimulators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	26/02/2023	26/02/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	26/02/2023	26/02/2023
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	26/02/2023	26/02/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	26/02/2023	26/02/2023
					CT2985	Active	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	26/02/2023	26/02/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	26/02/2023	26/02/2023
					CT3001	Implantable metal	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	26/02/2023	26/02/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	26/02/2023	26/02/2023
					CT3002	Implantable synthetic polymer	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	26/02/2023	26/02/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	26/02/2023	26/02/2023
					CT3006	Implantable inorganic material	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	26/02/2023	26/02/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	26/02/2023	26/02/2023
					CT314	Electrically powered	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	26/02/2023	26/02/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	26/02/2023	26/02/2023
					CT321	Long-term surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	26/02/2023	26/02/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	26/02/2023	26/02/2023
					CT449	Generators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	26/02/2023	26/02/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	26/02/2023	26/02/2023
					CT981	Single use	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	26/02/2023	26/02/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	26/02/2023	26/02/2023
					CT983	Surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	26/02/2023	26/02/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	26/02/2023	26/02/2023
13759	2023		47265	Dual-chamber implantable pacemaker, rate-responsive	CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem		
							F1901	Additional Surgery	A24	No Apparent Device Problem		
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem		
							F1901	Additional Surgery	A24	No Apparent Device Problem		
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem		
							F1901	Additional Surgery	A24	No Apparent Device Problem		
13760	2023		47265	Dual-chamber implantable pacemaker, rate-responsive	CT141	Electrodes/Leads and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	30/06/2023	30/06/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	30/06/2023	30/06/2023
					CT145	Cardiovascular devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	30/06/2023	30/06/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	30/06/2023	30/06/2023
					CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	30/06/2023	30/06/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	30/06/2023	30/06/2023
					CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	30/06/2023	30/06/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	30/06/2023	30/06/2023
					CT233	Surgical	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	30/06/2023	30/06/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	30/06/2023	30/06/2023
					CT2406	Implantable	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	30/06/2023	30/06/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	30/06/2023	30/06/2023
					CT2407	Active implantable	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	30/06/2023	30/06/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	30/06/2023	30/06/2023
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	30/06/2023	30/06/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	30/06/2023	30/06/2023
					CT2727	Active implant devices/ancillaries	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	30/06/2023	30/06/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	30/06/2023	30/06/2023
					CT2750	Implantable cardiac electrical stimulators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	30/06/2023	30/06/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	30/06/2023	30/06/2023
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	30/06/2023	30/06/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	30/06/2023	30/06/2023
					CT2985	Active	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	30/06/2023	30/06/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	30/06/2023	30/06/2023
					CT3001	Implantable metal	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	30/06/2023	30/06/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	30/06/2023	30/06/2023
					CT3002	Implantable synthetic polymer	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	30/06/2023	30/06/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	30/06/2023	30/06/2023
					CT3006	Implantable inorganic material	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	30/06/2023	30/06/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	30/06/2023	30/06/2023
					CT314	Electrically powered	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	30/06/2023	30/06/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	30/06/2023	30/06/2023
					CT321	Long-term surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	30/06/2023	30/06/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	30/06/2023	30/06/2023
					CT449	Generators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	30/06/2023	30/06/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	30/06/2023	30/06/2023
					CT981	Single use	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	30/06/2023	30/06/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	30/06/2023	30/06/2023
					CT983	Surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	30/06/2023	30/06/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	30/06/2023	30/06/2023
13761	2023		47265	Dual-chamber implantable pacemaker, rate-responsive	CT141	Electrodes/Leads and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	29/06/2023	29/06/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	29/06/2023	29/06/2023
					CT145	Cardiovascular devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	29/06/2023	29/06/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	29/06/2023	29/06/2023
					CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	29/06/2023	29/06/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	29/06/2023	29/06/2023
					CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	29/06/2023	29/06/2023

							F1901	Additional Surgery	A24	No Apparent Device Problem	29/06/2023	29/06/2023
					CT233	Surgical	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	29/06/2023	29/06/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	29/06/2023	29/06/2023
					CT2406	Implantable	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	29/06/2023	29/06/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	29/06/2023	29/06/2023
					CT2407	Active implantable	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	29/06/2023	29/06/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	29/06/2023	29/06/2023
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	29/06/2023	29/06/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	29/06/2023	29/06/2023
					CT2727	Active implant devices/ancillaries	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	29/06/2023	29/06/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	29/06/2023	29/06/2023
					CT2750	Implantable cardiac electrical stimulators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	29/06/2023	29/06/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	29/06/2023	29/06/2023
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	29/06/2023	29/06/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	29/06/2023	29/06/2023
					CT2985	Active	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	29/06/2023	29/06/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	29/06/2023	29/06/2023
					CT3001	Implantable metal	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	29/06/2023	29/06/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	29/06/2023	29/06/2023
					CT3002	Implantable synthetic polymer	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	29/06/2023	29/06/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	29/06/2023	29/06/2023
					CT3006	Implantable inorganic material	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	29/06/2023	29/06/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	29/06/2023	29/06/2023
					CT314	Electrically powered	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	29/06/2023	29/06/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	29/06/2023	29/06/2023
					CT321	Long-term surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	29/06/2023	29/06/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	29/06/2023	29/06/2023
					CT449	Generators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	29/06/2023	29/06/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	29/06/2023	29/06/2023
					CT981	Single use	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	29/06/2023	29/06/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	29/06/2023	29/06/2023
					CT983	Surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	29/06/2023	29/06/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	29/06/2023	29/06/2023
13762	2023	All other reportable incidents	47265	Dual-chamber implantable pacemaker, rate-responsive	CT141	Electrodes/Leads and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	19/07/2023	19/07/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	19/07/2023	19/07/2023
					CT145	Cardiovascular devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	19/07/2023	19/07/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	19/07/2023	19/07/2023
					CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	19/07/2023	19/07/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	19/07/2023	19/07/2023
					CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	19/07/2023	19/07/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	19/07/2023	19/07/2023
					CT233	Surgical	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	19/07/2023	19/07/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	19/07/2023	19/07/2023
					CT2406	Implantable	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	19/07/2023	19/07/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	19/07/2023	19/07/2023
					CT2407	Active implantable	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	19/07/2023	19/07/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	19/07/2023	19/07/2023
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	19/07/2023	19/07/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	19/07/2023	19/07/2023
					CT2727	Active implant devices/ancillaries	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	19/07/2023	19/07/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	19/07/2023	19/07/2023
					CT2750	Implantable cardiac electrical stimulators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	19/07/2023	19/07/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	19/07/2023	19/07/2023
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	19/07/2023	19/07/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	19/07/2023	19/07/2023
					CT2985	Active	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	19/07/2023	19/07/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	19/07/2023	19/07/2023
					CT3001	Implantable metal	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	19/07/2023	19/07/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	19/07/2023	19/07/2023
					CT3002	Implantable synthetic polymer	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	19/07/2023	19/07/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	19/07/2023	19/07/2023
					CT3006	Implantable inorganic material	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	19/07/2023	19/07/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	19/07/2023	19/07/2023
					CT314	Electrically powered	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	19/07/2023	19/07/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	19/07/2023	19/07/2023
					CT321	Long-term surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	19/07/2023	19/07/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	19/07/2023	19/07/2023
					CT449	Generators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	19/07/2023	19/07/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	19/07/2023	19/07/2023
					CT981	Single use	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	19/07/2023	19/07/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	19/07/2023	19/07/2023
					CT983	Surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	19/07/2023	19/07/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	19/07/2023	19/07/2023
13763	2023	All other reportable incidents	17882	Physiologic-monitoring defibrillation system	CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070803	Failure to Power Up		
							F27	Problem Identified During Non-Clinical Procedure	A070803	Failure to Power Up		
13764	2023	All other reportable incidents	47265	Dual-chamber implantable pacemaker, rate-responsive	CT141	Electrodes/Leads and associated devices	E0112	Dizziness	A070504	Premature Discharge of Battery	31/07/2023	31/07/2023
									A071204	No Pacing	31/07/2023	31/07/2023
							F1901	Additional Surgery	A070504	Premature Discharge of Battery	31/07/2023	31/07/2023
									A071204	No Pacing	31/07/2023	31/07/2023
					CT145	Cardiovascular devices	E0112	Dizziness	A070504	Premature Discharge of Battery	31/07/2023	31/07/2023
									A071204	No Pacing	31/07/2023	31/07/2023
							F1901	Additional Surgery	A070504	Premature Discharge of Battery	31/07/2023	31/07/2023
									A071204	No Pacing	31/07/2023	31/07/2023
					CT1684	Cardiac pacemakers and associated devices	E0112	Dizziness	A070504	Premature Discharge of Battery	31/07/2023	31/07/2023
									A071204	No Pacing	31/07/2023	31/07/2023
							F1901	Additional Surgery	A070504	Premature Discharge of Battery	31/07/2023	31/07/2023
									A071204	No Pacing	31/07/2023	31/07/2023
					CT1685	Cardiac pacemakers	E0112	Dizziness	A070504	Premature Discharge of Battery	31/07/2023	31/07/2023

									A071204	No Pacing	31/07/2023	31/07/2023
							F1901	Additional Surgery	A070504	Premature Discharge of Battery	31/07/2023	31/07/2023
									A071204	No Pacing	31/07/2023	31/07/2023
					CT233	Surgical	E0112	Dizziness	A070504	Premature Discharge of Battery	31/07/2023	31/07/2023
									A071204	No Pacing	31/07/2023	31/07/2023
							F1901	Additional Surgery	A070504	Premature Discharge of Battery	31/07/2023	31/07/2023
									A071204	No Pacing	31/07/2023	31/07/2023
					CT2406	Implantable	E0112	Dizziness	A070504	Premature Discharge of Battery	31/07/2023	31/07/2023
									A071204	No Pacing	31/07/2023	31/07/2023
							F1901	Additional Surgery	A070504	Premature Discharge of Battery	31/07/2023	31/07/2023
									A071204	No Pacing	31/07/2023	31/07/2023
					CT2407	Active implantable	E0112	Dizziness	A070504	Premature Discharge of Battery	31/07/2023	31/07/2023
									A071204	No Pacing	31/07/2023	31/07/2023
							F1901	Additional Surgery	A070504	Premature Discharge of Battery	31/07/2023	31/07/2023
									A071204	No Pacing	31/07/2023	31/07/2023
					CT257	Implantable cardiac pacemakers	E0112	Dizziness	A070504	Premature Discharge of Battery	31/07/2023	31/07/2023
									A071204	No Pacing	31/07/2023	31/07/2023
							F1901	Additional Surgery	A070504	Premature Discharge of Battery	31/07/2023	31/07/2023
									A071204	No Pacing	31/07/2023	31/07/2023
					CT2727	Active implant devices/ancillaries	E0112	Dizziness	A070504	Premature Discharge of Battery	31/07/2023	31/07/2023
									A071204	No Pacing	31/07/2023	31/07/2023
							F1901	Additional Surgery	A070504	Premature Discharge of Battery	31/07/2023	31/07/2023
									A071204	No Pacing	31/07/2023	31/07/2023
					CT2750	Implantable cardiac electrical stimulators and associated devices	E0112	Dizziness	A070504	Premature Discharge of Battery	31/07/2023	31/07/2023
									A071204	No Pacing	31/07/2023	31/07/2023
							F1901	Additional Surgery	A070504	Premature Discharge of Battery	31/07/2023	31/07/2023
									A071204	No Pacing	31/07/2023	31/07/2023
					CT2866	Implantable cardiac electrical stimulators	E0112	Dizziness	A070504	Premature Discharge of Battery	31/07/2023	31/07/2023
									A071204	No Pacing	31/07/2023	31/07/2023
							F1901	Additional Surgery	A070504	Premature Discharge of Battery	31/07/2023	31/07/2023
									A071204	No Pacing	31/07/2023	31/07/2023
					CT2985	Active	E0112	Dizziness	A070504	Premature Discharge of Battery	31/07/2023	31/07/2023
									A071204	No Pacing	31/07/2023	31/07/2023
							F1901	Additional Surgery	A070504	Premature Discharge of Battery	31/07/2023	31/07/2023
									A071204	No Pacing	31/07/2023	31/07/2023
					CT3001	Implantable metal	E0112	Dizziness	A070504	Premature Discharge of Battery	31/07/2023	31/07/2023
									A071204	No Pacing	31/07/2023	31/07/2023
							F1901	Additional Surgery	A070504	Premature Discharge of Battery	31/07/2023	31/07/2023
									A071204	No Pacing	31/07/2023	31/07/2023
					CT3002	Implantable synthetic polymer	E0112	Dizziness	A070504	Premature Discharge of Battery	31/07/2023	31/07/2023
									A071204	No Pacing	31/07/2023	31/07/2023
							F1901	Additional Surgery	A070504	Premature Discharge of Battery	31/07/2023	31/07/2023
									A071204	No Pacing	31/07/2023	31/07/2023
					CT3006	Implantable inorganic material	E0112	Dizziness	A070504	Premature Discharge of Battery	31/07/2023	31/07/2023
									A071204	No Pacing	31/07/2023	31/07/2023
							F1901	Additional Surgery	A070504	Premature Discharge of Battery	31/07/2023	31/07/2023
									A071204	No Pacing	31/07/2023	31/07/2023
					CT314	Electrically powered	E0112	Dizziness	A070504	Premature Discharge of Battery	31/07/2023	31/07/2023
									A071204	No Pacing	31/07/2023	31/07/2023
							F1901	Additional Surgery	A070504	Premature Discharge of Battery	31/07/2023	31/07/2023
									A071204	No Pacing	31/07/2023	31/07/2023
					CT321	Long-term surgical invasive	E0112	Dizziness	A070504	Premature Discharge of Battery	31/07/2023	31/07/2023
									A071204	No Pacing	31/07/2023	31/07/2023
							F1901	Additional Surgery	A070504	Premature Discharge of Battery	31/07/2023	31/07/2023
									A071204	No Pacing	31/07/2023	31/07/2023
					CT449	Generators and associated devices	E0112	Dizziness	A070504	Premature Discharge of Battery	31/07/2023	31/07/2023
									A071204	No Pacing	31/07/2023	31/07/2023
							F1901	Additional Surgery	A070504	Premature Discharge of Battery	31/07/2023	31/07/2023
									A071204	No Pacing	31/07/2023	31/07/2023
					CT981	Single use	E0112	Dizziness	A070504	Premature Discharge of Battery	31/07/2023	31/07/2023
									A071204	No Pacing	31/07/2023	31/07/2023
							F1901	Additional Surgery	A070504	Premature Discharge of Battery	31/07/2023	31/07/2023
									A071204	No Pacing	31/07/2023	31/07/2023
					CT983	Surgical invasive	E0112	Dizziness	A070504	Premature Discharge of Battery	31/07/2023	31/07/2023
									A071204	No Pacing	31/07/2023	31/07/2023
							F1901	Additional Surgery	A070504	Premature Discharge of Battery	31/07/2023	31/07/2023
									A071204	No Pacing	31/07/2023	31/07/2023
13765	2023		47265	Dual-chamber implantable pacemaker, rate-responsive	CT141	Electrodes/Leads and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	24/07/2023	24/07/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	24/07/2023	24/07/2023
					CT145	Cardiovascular devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	24/07/2023	24/07/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	24/07/2023	24/07/2023
					CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	24/07/2023	24/07/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	24/07/2023	24/07/2023
					CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	24/07/2023	24/07/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	24/07/2023	24/07/2023
					CT233	Surgical	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	24/07/2023	24/07/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	24/07/2023	24/07/2023
					CT2406	Implantable	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	24/07/2023	24/07/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	24/07/2023	24/07/2023
					CT2407	Active implantable	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	24/07/2023	24/07/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	24/07/2023	24/07/2023
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	24/07/2023	24/07/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	24/07/2023	24/07/2023
					CT2727	Active implant devices/ancillaries	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	24/07/2023	24/07/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	24/07/2023	24/07/2023
					CT2750	Implantable cardiac electrical stimulators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	24/07/2023	24/07/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	24/07/2023	24/07/2023
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	24/07/2023	24/07/2023

							F1901	Additional Surgery	A24	No Apparent Device Problem	24/07/2023	24/07/2023
					CT2985	Active	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	24/07/2023	24/07/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	24/07/2023	24/07/2023
					CT3001	Implantable metal	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	24/07/2023	24/07/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	24/07/2023	24/07/2023
					CT3002	Implantable synthetic polymer	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	24/07/2023	24/07/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	24/07/2023	24/07/2023
					CT3006	Implantable inorganic material	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	24/07/2023	24/07/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	24/07/2023	24/07/2023
					CT314	Electrically powered	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	24/07/2023	24/07/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	24/07/2023	24/07/2023
					CT321	Long-term surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	24/07/2023	24/07/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	24/07/2023	24/07/2023
					CT449	Generators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	24/07/2023	24/07/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	24/07/2023	24/07/2023
					CT981	Single use	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	24/07/2023	24/07/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	24/07/2023	24/07/2023
					CT983	Surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	24/07/2023	24/07/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	24/07/2023	24/07/2023
13766	2023		47265	Dual-chamber implantable pacemaker, rate-responsive	CT141	Electrodes/Leads and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	20/07/2023	20/07/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	20/07/2023	20/07/2023
					CT145	Cardiovascular devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	20/07/2023	20/07/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	20/07/2023	20/07/2023
					CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	20/07/2023	20/07/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	20/07/2023	20/07/2023
					CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	20/07/2023	20/07/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	20/07/2023	20/07/2023
					CT233	Surgical	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	20/07/2023	20/07/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	20/07/2023	20/07/2023
					CT2406	Implantable	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	20/07/2023	20/07/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	20/07/2023	20/07/2023
					CT2407	Active implantable	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	20/07/2023	20/07/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	20/07/2023	20/07/2023
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	20/07/2023	20/07/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	20/07/2023	20/07/2023
					CT2727	Active implant devices/ancillaries	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	20/07/2023	20/07/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	20/07/2023	20/07/2023
					CT2750	Implantable cardiac electrical stimulators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	20/07/2023	20/07/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	20/07/2023	20/07/2023
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	20/07/2023	20/07/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	20/07/2023	20/07/2023
					CT2985	Active	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	20/07/2023	20/07/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	20/07/2023	20/07/2023
					CT3001	Implantable metal	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	20/07/2023	20/07/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	20/07/2023	20/07/2023
					CT3002	Implantable synthetic polymer	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	20/07/2023	20/07/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	20/07/2023	20/07/2023
					CT3006	Implantable inorganic material	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	20/07/2023	20/07/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	20/07/2023	20/07/2023
					CT314	Electrically powered	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	20/07/2023	20/07/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	20/07/2023	20/07/2023
					CT321	Long-term surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	20/07/2023	20/07/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	20/07/2023	20/07/2023
					CT449	Generators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	20/07/2023	20/07/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	20/07/2023	20/07/2023
					CT981	Single use	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	20/07/2023	20/07/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	20/07/2023	20/07/2023
					CT983	Surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	20/07/2023	20/07/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	20/07/2023	20/07/2023
13767	2023	All other reportable incidents	35223	Endocardial/interventricular septal pacing lead	CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A0401	Break		
									A070909	Oversensing		
									A090801	Signal Artifact/Noise		
							F26	No Health Consequences or Impact	A0401	Break		
									A070909	Oversensing		
									A090801	Signal Artifact/Noise		
13768	2023		47265	Dual-chamber implantable pacemaker, rate-responsive	CT141	Electrodes/Leads and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem		
							F1901	Additional Surgery	A24	No Apparent Device Problem		
					CT145	Cardiovascular devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem		
							F1901	Additional Surgery	A24	No Apparent Device Problem		
					CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem		
							F1901	Additional Surgery	A24	No Apparent Device Problem		
					CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem		
							F1901	Additional Surgery	A24	No Apparent Device Problem		
					CT233	Surgical	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem		
							F1901	Additional Surgery	A24	No Apparent Device Problem		
					CT2406	Implantable	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem		
							F1901	Additional Surgery	A24	No Apparent Device Problem		
					CT2407	Active implantable	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem		
							F1901	Additional Surgery	A24	No Apparent Device Problem		
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem		
							F1901	Additional Surgery	A24	No Apparent Device Problem		
					CT2727	Active implant devices/ancillaries	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem		
							F1901	Additional Surgery	A24	No Apparent Device Problem		
					CT2750	Implantable cardiac electrical stimulators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem		
							F1901	Additional Surgery	A24	No Apparent Device Problem		
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem		
							F1901	Additional Surgery	A24	No Apparent Device Problem		
					CT2985	Active	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem		

							F1901	Additional Surgery	A24	No Apparent Device Problem		
					CT3001	Implantable metal	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem		
							F1901	Additional Surgery	A24	No Apparent Device Problem		
					CT3002	Implantable synthetic polymer	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem		
							F1901	Additional Surgery	A24	No Apparent Device Problem		
					CT3006	Implantable inorganic material	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem		
							F1901	Additional Surgery	A24	No Apparent Device Problem		
					CT314	Electrically powered	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem		
							F1901	Additional Surgery	A24	No Apparent Device Problem		
					CT321	Long-term surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem		
							F1901	Additional Surgery	A24	No Apparent Device Problem		
					CT449	Generators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem		
							F1901	Additional Surgery	A24	No Apparent Device Problem		
					CT981	Single use	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem		
							F1901	Additional Surgery	A24	No Apparent Device Problem		
					CT983	Surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem		
							F1901	Additional Surgery	A24	No Apparent Device Problem		
13769	2023	All other reportable incidents	17882	Physiologic-monitoring defibrillation system	CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A040101	Fracture		
									A07	Electrical /Electronic Property Problem		
							F27	Problem Identified During Non-Clinical Procedure	A040101	Fracture		
									A07	Electrical /Electronic Property Problem		
13770	2023	All other reportable incidents	47265	Dual-chamber implantable pacemaker, rate-responsive	CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	26/02/2023	26/02/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	26/02/2023	26/02/2023
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	26/02/2023	26/02/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	26/02/2023	26/02/2023
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	26/02/2023	26/02/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	26/02/2023	26/02/2023
13771	2023	All other reportable incidents	47265	Dual-chamber implantable pacemaker, rate-responsive	CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A0203	Defective Device	06/01/2023	06/01/2023
									A070909	Oversensing	06/01/2023	06/01/2023
							F1901	Additional Surgery	A0203	Defective Device	06/01/2023	06/01/2023
									A070909	Oversensing	06/01/2023	06/01/2023
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A0203	Defective Device	06/01/2023	06/01/2023
									A070909	Oversensing	06/01/2023	06/01/2023
							F1901	Additional Surgery	A0203	Defective Device	06/01/2023	06/01/2023
									A070909	Oversensing	06/01/2023	06/01/2023
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A0203	Defective Device	06/01/2023	06/01/2023
									A070909	Oversensing	06/01/2023	06/01/2023
							F1901	Additional Surgery	A0203	Defective Device	06/01/2023	06/01/2023
									A070909	Oversensing	06/01/2023	06/01/2023
13772	2023		47265	Dual-chamber implantable pacemaker, rate-responsive	CT141	Electrodes/Leads and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	16/03/2023	16/03/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	16/03/2023	16/03/2023
					CT145	Cardiovascular devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	16/03/2023	16/03/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	16/03/2023	16/03/2023
					CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	16/03/2023	16/03/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	16/03/2023	16/03/2023
					CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	16/03/2023	16/03/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	16/03/2023	16/03/2023
					CT233	Surgical	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	16/03/2023	16/03/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	16/03/2023	16/03/2023
					CT2406	Implantable	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	16/03/2023	16/03/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	16/03/2023	16/03/2023
					CT2407	Active implantable	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	16/03/2023	16/03/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	16/03/2023	16/03/2023
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	16/03/2023	16/03/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	16/03/2023	16/03/2023
					CT2727	Active implant devices/ancillaries	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	16/03/2023	16/03/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	16/03/2023	16/03/2023
					CT2750	Implantable cardiac electrical stimulators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	16/03/2023	16/03/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	16/03/2023	16/03/2023
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	16/03/2023	16/03/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	16/03/2023	16/03/2023
					CT2985	Active	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	16/03/2023	16/03/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	16/03/2023	16/03/2023
					CT3001	Implantable metal	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	16/03/2023	16/03/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	16/03/2023	16/03/2023
					CT3002	Implantable synthetic polymer	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	16/03/2023	16/03/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	16/03/2023	16/03/2023
					CT3006	Implantable inorganic material	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	16/03/2023	16/03/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	16/03/2023	16/03/2023
					CT314	Electrically powered	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	16/03/2023	16/03/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	16/03/2023	16/03/2023
					CT321	Long-term surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	16/03/2023	16/03/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	16/03/2023	16/03/2023
					CT449	Generators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	16/03/2023	16/03/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	16/03/2023	16/03/2023
					CT981	Single use	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	16/03/2023	16/03/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	16/03/2023	16/03/2023
					CT983	Surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	16/03/2023	16/03/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	16/03/2023	16/03/2023
13773	2023	All other reportable incidents	47270	Cardiac resynchronization therapy implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A160601	Premature Elective Replacement Indicator	04/08/2023	04/08/2023
							F1903	Device Explantation	A160601	Premature Elective Replacement Indicator	04/08/2023	04/08/2023
13774	2023		47265	Dual-chamber implantable pacemaker, rate-responsive	CT141	Electrodes/Leads and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A0203	Defective Device		
							F08	Hospitalization or Prolonged Hospitalization	A0203	Defective Device		
					CT145	Cardiovascular devices	E2403	No Clinical Signs, Symptoms or Conditions	A0203	Defective Device		
							F08	Hospitalization or Prolonged Hospitalization	A0203	Defective Device		
					CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A0203	Defective Device		
							F08	Hospitalization or Prolonged Hospitalization	A0203	Defective Device		
					CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A0203	Defective Device		

							F08	Hospitalization or Prolonged Hospitalization	A0203	Defective Device			
					CT233	Surgical	E2403	No Clinical Signs, Symptoms or Conditions	A0203	Defective Device			
							F08	Hospitalization or Prolonged Hospitalization	A0203	Defective Device			
					CT2406	Implantable	E2403	No Clinical Signs, Symptoms or Conditions	A0203	Defective Device			
							F08	Hospitalization or Prolonged Hospitalization	A0203	Defective Device			
					CT2407	Active implantable	E2403	No Clinical Signs, Symptoms or Conditions	A0203	Defective Device			
							F08	Hospitalization or Prolonged Hospitalization	A0203	Defective Device			
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A0203	Defective Device			
							F08	Hospitalization or Prolonged Hospitalization	A0203	Defective Device			
					CT2727	Active implant devices/ancillaries	E2403	No Clinical Signs, Symptoms or Conditions	A0203	Defective Device			
							F08	Hospitalization or Prolonged Hospitalization	A0203	Defective Device			
					CT2750	Implantable cardiac electrical stimulators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A0203	Defective Device			
							F08	Hospitalization or Prolonged Hospitalization	A0203	Defective Device			
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A0203	Defective Device			
							F08	Hospitalization or Prolonged Hospitalization	A0203	Defective Device			
					CT2985	Active	E2403	No Clinical Signs, Symptoms or Conditions	A0203	Defective Device			
							F08	Hospitalization or Prolonged Hospitalization	A0203	Defective Device			
					CT3001	Implantable metal	E2403	No Clinical Signs, Symptoms or Conditions	A0203	Defective Device			
							F08	Hospitalization or Prolonged Hospitalization	A0203	Defective Device			
					CT3002	Implantable synthetic polymer	E2403	No Clinical Signs, Symptoms or Conditions	A0203	Defective Device			
							F08	Hospitalization or Prolonged Hospitalization	A0203	Defective Device			
					CT3006	Implantable inorganic material	E2403	No Clinical Signs, Symptoms or Conditions	A0203	Defective Device			
							F08	Hospitalization or Prolonged Hospitalization	A0203	Defective Device			
					CT314	Electrically powered	E2403	No Clinical Signs, Symptoms or Conditions	A0203	Defective Device			
							F08	Hospitalization or Prolonged Hospitalization	A0203	Defective Device			
					CT321	Long-term surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A0203	Defective Device			
							F08	Hospitalization or Prolonged Hospitalization	A0203	Defective Device			
					CT449	Generators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A0203	Defective Device			
							F08	Hospitalization or Prolonged Hospitalization	A0203	Defective Device			
					CT981	Single use	E2403	No Clinical Signs, Symptoms or Conditions	A0203	Defective Device			
							F08	Hospitalization or Prolonged Hospitalization	A0203	Defective Device			
					CT983	Surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A0203	Defective Device			
							F08	Hospitalization or Prolonged Hospitalization	A0203	Defective Device			
13775	2023	All other reportable incidents	35223	Endocardial/interventricular septal pacing lead	CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070909	Oversensing			
									A090801	Signal Artifact/Noise			
							F08	Hospitalization or Prolonged Hospitalization	A070909	Oversensing			
									A090801	Signal Artifact/Noise			
							F12	Serious Injury/ Illness/ Impairment	A070909	Oversensing			
									A090801	Signal Artifact/Noise			
							F23	Unexpected Medical Intervention	A070909	Oversensing			
									A090801	Signal Artifact/Noise			
13776	2023	All other reportable incidents	47265	Dual-chamber implantable pacemaker, rate-responsive	CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	16/05/2023	16/05/2023	
							F1901	Additional Surgery	A24	No Apparent Device Problem	16/05/2023	16/05/2023	
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	16/05/2023	16/05/2023	
							F1901	Additional Surgery	A24	No Apparent Device Problem	16/05/2023	16/05/2023	
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	16/05/2023	16/05/2023	
							F1901	Additional Surgery	A24	No Apparent Device Problem	16/05/2023	16/05/2023	
13777	2023	All other reportable incidents	47265	Dual-chamber implantable pacemaker, rate-responsive	CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	27/02/2023	27/02/2023	
							F1901	Additional Surgery	A24	No Apparent Device Problem	27/02/2023	27/02/2023	
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	27/02/2023	27/02/2023	
							F1901	Additional Surgery	A24	No Apparent Device Problem	27/02/2023	27/02/2023	
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	27/02/2023	27/02/2023	
							F1901	Additional Surgery	A24	No Apparent Device Problem	27/02/2023	27/02/2023	
13778	2023		47265	Dual-chamber implantable pacemaker, rate-responsive	CT141	Electrodes/Leads and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	05/02/2023	05/02/2023	
							F1901	Additional Surgery	A24	No Apparent Device Problem	05/02/2023	05/02/2023	
					CT145	Cardiovascular devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	05/02/2023	05/02/2023	
							F1901	Additional Surgery	A24	No Apparent Device Problem	05/02/2023	05/02/2023	
					CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	05/02/2023	05/02/2023	
							F1901	Additional Surgery	A24	No Apparent Device Problem	05/02/2023	05/02/2023	
					CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	05/02/2023	05/02/2023	
							F1901	Additional Surgery	A24	No Apparent Device Problem	05/02/2023	05/02/2023	
					CT233	Surgical	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	05/02/2023	05/02/2023	
							F1901	Additional Surgery	A24	No Apparent Device Problem	05/02/2023	05/02/2023	
					CT2406	Implantable	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	05/02/2023	05/02/2023	
							F1901	Additional Surgery	A24	No Apparent Device Problem	05/02/2023	05/02/2023	
					CT2407	Active implantable	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	05/02/2023	05/02/2023	
							F1901	Additional Surgery	A24	No Apparent Device Problem	05/02/2023	05/02/2023	
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	05/02/2023	05/02/2023	
							F1901	Additional Surgery	A24	No Apparent Device Problem	05/02/2023	05/02/2023	
					CT2727	Active implant devices/ancillaries	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	05/02/2023	05/02/2023	
							F1901	Additional Surgery	A24	No Apparent Device Problem	05/02/2023	05/02/2023	
					CT2750	Implantable cardiac electrical stimulators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	05/02/2023	05/02/2023	
							F1901	Additional Surgery	A24	No Apparent Device Problem	05/02/2023	05/02/2023	
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	05/02/2023	05/02/2023	
							F1901	Additional Surgery	A24	No Apparent Device Problem	05/02/2023	05/02/2023	
					CT2985	Active	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	05/02/2023	05/02/2023	
							F1901	Additional Surgery	A24	No Apparent Device Problem	05/02/2023	05/02/2023	
					CT3001	Implantable metal	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	05/02/2023	05/02/2023	
							F1901	Additional Surgery	A24	No Apparent Device Problem	05/02/2023	05/02/2023	
					CT3002	Implantable synthetic polymer	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	05/02/2023	05/02/2023	
							F1901	Additional Surgery	A24	No Apparent Device Problem	05/02/2023	05/02/2023	
					CT3006	Implantable inorganic material	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	05/02/2023	05/02/2023	
							F1901	Additional Surgery	A24	No Apparent Device Problem	05/02/2023	05/02/2023	
					CT314	Electrically powered	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	05/02/2023	05/02/2023	
							F1901	Additional Surgery	A24	No Apparent Device Problem	05/02/2023	05/02/2023	
					CT321	Long-term surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	05/02/2023	05/02/2023	
							F1901	Additional Surgery	A24	No Apparent Device Problem	05/02/2023	05/02/2023	
					CT449	Generators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	05/02/2023	05/02/2023	

						F1901	Additional Surgery	A24	No Apparent Device Problem	05/02/2023	05/02/2023
				CT981	Single use	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	05/02/2023	05/02/2023
						F1901	Additional Surgery	A24	No Apparent Device Problem	05/02/2023	05/02/2023
				CT983	Surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	05/02/2023	05/02/2023
						F1901	Additional Surgery	A24	No Apparent Device Problem	05/02/2023	05/02/2023
13779	2023	All other reportable incidents	47265	Dual-chamber implantable pacemaker, rate-responsive	CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	06/03/2023	06/03/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	06/03/2023	06/03/2023
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	06/03/2023	06/03/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	06/03/2023	06/03/2023
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	06/03/2023	06/03/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	06/03/2023	06/03/2023
13780	2023		47265	Dual-chamber implantable pacemaker, rate-responsive	CT141	Electrodes/Leads and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	17/03/2023	17/03/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	17/03/2023	17/03/2023
					CT145	Cardiovascular devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	17/03/2023	17/03/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	17/03/2023	17/03/2023
					CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	17/03/2023	17/03/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	17/03/2023	17/03/2023
					CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	17/03/2023	17/03/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	17/03/2023	17/03/2023
					CT233	Surgical	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	17/03/2023	17/03/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	17/03/2023	17/03/2023
					CT2406	Implantable	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	17/03/2023	17/03/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	17/03/2023	17/03/2023
					CT2407	Active implantable	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	17/03/2023	17/03/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	17/03/2023	17/03/2023
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	17/03/2023	17/03/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	17/03/2023	17/03/2023
					CT2727	Active implant devices/ancillaries	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	17/03/2023	17/03/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	17/03/2023	17/03/2023
					CT2750	Implantable cardiac electrical stimulators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	17/03/2023	17/03/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	17/03/2023	17/03/2023
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	17/03/2023	17/03/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	17/03/2023	17/03/2023
					CT2985	Active	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	17/03/2023	17/03/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	17/03/2023	17/03/2023
					CT3001	Implantable metal	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	17/03/2023	17/03/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	17/03/2023	17/03/2023
					CT3002	Implantable synthetic polymer	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	17/03/2023	17/03/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	17/03/2023	17/03/2023
					CT3006	Implantable inorganic material	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	17/03/2023	17/03/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	17/03/2023	17/03/2023
					CT314	Electrically powered	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	17/03/2023	17/03/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	17/03/2023	17/03/2023
					CT321	Long-term surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	17/03/2023	17/03/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	17/03/2023	17/03/2023
					CT449	Generators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	17/03/2023	17/03/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	17/03/2023	17/03/2023
					CT981	Single use	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	17/03/2023	17/03/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	17/03/2023	17/03/2023
					CT983	Surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	17/03/2023	17/03/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	17/03/2023	17/03/2023
13781	2023	All other reportable incidents	47265	Dual-chamber implantable pacemaker, rate-responsive	CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	03/04/2023	03/04/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	03/04/2023	03/04/2023
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	03/04/2023	03/04/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	03/04/2023	03/04/2023
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	03/04/2023	03/04/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	03/04/2023	03/04/2023
13782	2023		47265	Dual-chamber implantable pacemaker, rate-responsive	CT141	Electrodes/Leads and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	26/02/2023	26/02/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	26/02/2023	26/02/2023
					CT145	Cardiovascular devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	26/02/2023	26/02/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	26/02/2023	26/02/2023
					CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	26/02/2023	26/02/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	26/02/2023	26/02/2023
					CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	26/02/2023	26/02/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	26/02/2023	26/02/2023
					CT233	Surgical	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	26/02/2023	26/02/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	26/02/2023	26/02/2023
					CT2406	Implantable	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	26/02/2023	26/02/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	26/02/2023	26/02/2023
					CT2407	Active implantable	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	26/02/2023	26/02/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	26/02/2023	26/02/2023
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	26/02/2023	26/02/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	26/02/2023	26/02/2023
					CT2727	Active implant devices/ancillaries	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	26/02/2023	26/02/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	26/02/2023	26/02/2023
					CT2750	Implantable cardiac electrical stimulators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	26/02/2023	26/02/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	26/02/2023	26/02/2023
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	26/02/2023	26/02/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	26/02/2023	26/02/2023
					CT2985	Active	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	26/02/2023	26/02/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	26/02/2023	26/02/2023
					CT3001	Implantable metal	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	26/02/2023	26/02/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	26/02/2023	26/02/2023
					CT3002	Implantable synthetic polymer	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	26/02/2023	26/02/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	26/02/2023	26/02/2023
					CT3006	Implantable inorganic material	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	26/02/2023	26/02/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	26/02/2023	26/02/2023
					CT314	Electrically powered	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	26/02/2023	26/02/2023

							F1901	Additional Surgery	A24	No Apparent Device Problem	26/02/2023	26/02/2023
					CT321	Long-term surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	26/02/2023	26/02/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	26/02/2023	26/02/2023
					CT449	Generators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	26/02/2023	26/02/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	26/02/2023	26/02/2023
					CT981	Single use	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	26/02/2023	26/02/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	26/02/2023	26/02/2023
					CT983	Surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	26/02/2023	26/02/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	26/02/2023	26/02/2023
13783	2023		47265	Dual-chamber implantable pacemaker, rate-responsive	CT141	Electrodes/Leads and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	23/06/2023	23/06/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	23/06/2023	23/06/2023
					CT145	Cardiovascular devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	23/06/2023	23/06/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	23/06/2023	23/06/2023
					CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	23/06/2023	23/06/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	23/06/2023	23/06/2023
					CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	23/06/2023	23/06/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	23/06/2023	23/06/2023
					CT233	Surgical	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	23/06/2023	23/06/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	23/06/2023	23/06/2023
					CT2406	Implantable	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	23/06/2023	23/06/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	23/06/2023	23/06/2023
					CT2407	Active implantable	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	23/06/2023	23/06/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	23/06/2023	23/06/2023
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	23/06/2023	23/06/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	23/06/2023	23/06/2023
					CT2727	Active implant devices/ancillaries	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	23/06/2023	23/06/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	23/06/2023	23/06/2023
					CT2750	Implantable cardiac electrical stimulators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	23/06/2023	23/06/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	23/06/2023	23/06/2023
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	23/06/2023	23/06/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	23/06/2023	23/06/2023
					CT2985	Active	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	23/06/2023	23/06/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	23/06/2023	23/06/2023
					CT3001	Implantable metal	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	23/06/2023	23/06/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	23/06/2023	23/06/2023
					CT3002	Implantable synthetic polymer	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	23/06/2023	23/06/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	23/06/2023	23/06/2023
					CT3006	Implantable inorganic material	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	23/06/2023	23/06/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	23/06/2023	23/06/2023
					CT314	Electrically powered	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	23/06/2023	23/06/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	23/06/2023	23/06/2023
					CT321	Long-term surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	23/06/2023	23/06/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	23/06/2023	23/06/2023
					CT449	Generators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	23/06/2023	23/06/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	23/06/2023	23/06/2023
					CT981	Single use	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	23/06/2023	23/06/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	23/06/2023	23/06/2023
					CT983	Surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	23/06/2023	23/06/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	23/06/2023	23/06/2023
13784	2023		47265	Dual-chamber implantable pacemaker, rate-responsive	CT141	Electrodes/Leads and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	30/06/2023	30/06/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	30/06/2023	30/06/2023
					CT145	Cardiovascular devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	30/06/2023	30/06/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	30/06/2023	30/06/2023
					CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	30/06/2023	30/06/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	30/06/2023	30/06/2023
					CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	30/06/2023	30/06/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	30/06/2023	30/06/2023
					CT233	Surgical	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	30/06/2023	30/06/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	30/06/2023	30/06/2023
					CT2406	Implantable	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	30/06/2023	30/06/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	30/06/2023	30/06/2023
					CT2407	Active implantable	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	30/06/2023	30/06/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	30/06/2023	30/06/2023
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	30/06/2023	30/06/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	30/06/2023	30/06/2023
					CT2727	Active implant devices/ancillaries	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	30/06/2023	30/06/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	30/06/2023	30/06/2023
					CT2750	Implantable cardiac electrical stimulators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	30/06/2023	30/06/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	30/06/2023	30/06/2023
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	30/06/2023	30/06/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	30/06/2023	30/06/2023
					CT2985	Active	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	30/06/2023	30/06/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	30/06/2023	30/06/2023
					CT3001	Implantable metal	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	30/06/2023	30/06/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	30/06/2023	30/06/2023
					CT3002	Implantable synthetic polymer	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	30/06/2023	30/06/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	30/06/2023	30/06/2023
					CT3006	Implantable inorganic material	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	30/06/2023	30/06/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	30/06/2023	30/06/2023
					CT314	Electrically powered	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	30/06/2023	30/06/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	30/06/2023	30/06/2023
					CT321	Long-term surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	30/06/2023	30/06/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	30/06/2023	30/06/2023
					CT449	Generators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	30/06/2023	30/06/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	30/06/2023	30/06/2023
					CT981	Single use	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	30/06/2023	30/06/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	30/06/2023	30/06/2023
					CT983	Surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	30/06/2023	30/06/2023

							F1901	Additional Surgery	A24	No Apparent Device Problem	30/06/2023	30/06/2023
13785	2023	All other reportable incidents	17882	Physiologic-monitoring defibrillation system	CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A0719	Unexpected Shutdown		
							F26	No Health Consequences or Impact	A0719	Unexpected Shutdown		
13786	2023	All other reportable incidents	46803	Cardiac device introducer	CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A040609	Material Twisted/Bent		
									A0509	Physical Resistance/Sticking		
							F1908	Prolonged Surgery	A040609	Material Twisted/Bent		
									A0509	Physical Resistance/Sticking		
13787	2023	All other reportable incidents	63188	Left-ventricular wireless pacing system sensor/transmitter	CT141	Electrodes/Leads and associated devices	E060109	Tachycardia	A0708	Power Problem		
							F05	Delay to Treatment/ Therapy	A0708	Power Problem		
					CT145	Cardiovascular devices	E060109	Tachycardia	A0708	Power Problem		
							F05	Delay to Treatment/ Therapy	A0708	Power Problem		
					CT1577	Wireless monitoring system transmitters	E060109	Tachycardia	A0708	Power Problem		
							F05	Delay to Treatment/ Therapy	A0708	Power Problem		
					CT1684	Cardiac pacemakers and associated devices	E060109	Tachycardia	A0708	Power Problem		
							F05	Delay to Treatment/ Therapy	A0708	Power Problem		
					CT2317	Transmitters/Receivers/Transceivers	E060109	Tachycardia	A0708	Power Problem		
							F05	Delay to Treatment/ Therapy	A0708	Power Problem		
					CT233	Surgical	E060109	Tachycardia	A0708	Power Problem		
							F05	Delay to Treatment/ Therapy	A0708	Power Problem		
					CT2406	Implantable	E060109	Tachycardia	A0708	Power Problem		
							F05	Delay to Treatment/ Therapy	A0708	Power Problem		
					CT2407	Active implantable	E060109	Tachycardia	A0708	Power Problem		
							F05	Delay to Treatment/ Therapy	A0708	Power Problem		
					CT255	Transducers and associated devices	E060109	Tachycardia	A0708	Power Problem		
							F05	Delay to Treatment/ Therapy	A0708	Power Problem		
					CT2726	Physiological monitoring/measuring devices	E060109	Tachycardia	A0708	Power Problem		
							F05	Delay to Treatment/ Therapy	A0708	Power Problem		
					CT2727	Active implant devices/ancillaries	E060109	Tachycardia	A0708	Power Problem		
							F05	Delay to Treatment/ Therapy	A0708	Power Problem		
					CT2750	Implantable cardiac electrical stimulators and associated devices	E060109	Tachycardia	A0708	Power Problem		
							F05	Delay to Treatment/ Therapy	A0708	Power Problem		
					CT2885	Physiological monitoring/measuring system transmitters/receivers/transceivers	E060109	Tachycardia	A0708	Power Problem		
							F05	Delay to Treatment/ Therapy	A0708	Power Problem		
					CT2985	Active	E060109	Tachycardia	A0708	Power Problem		
							F05	Delay to Treatment/ Therapy	A0708	Power Problem		
					CT3001	Implantable metal	E060109	Tachycardia	A0708	Power Problem		
							F05	Delay to Treatment/ Therapy	A0708	Power Problem		
					CT3002	Implantable synthetic polymer	E060109	Tachycardia	A0708	Power Problem		
							F05	Delay to Treatment/ Therapy	A0708	Power Problem		
					CT3006	Implantable inorganic material	E060109	Tachycardia	A0708	Power Problem		
							F05	Delay to Treatment/ Therapy	A0708	Power Problem		
					CT314	Electrically powered	E060109	Tachycardia	A0708	Power Problem		
							F05	Delay to Treatment/ Therapy	A0708	Power Problem		
					CT321	Long-term surgical invasive	E060109	Tachycardia	A0708	Power Problem		
							F05	Delay to Treatment/ Therapy	A0708	Power Problem		
					CT384	Sensors and associated devices	E060109	Tachycardia	A0708	Power Problem		
							F05	Delay to Treatment/ Therapy	A0708	Power Problem		
					CT449	Generators and associated devices	E060109	Tachycardia	A0708	Power Problem		
							F05	Delay to Treatment/ Therapy	A0708	Power Problem		
					CT981	Single use	E060109	Tachycardia	A0708	Power Problem		
							F05	Delay to Treatment/ Therapy	A0708	Power Problem		
					CT983	Surgical invasive	E060109	Tachycardia	A0708	Power Problem		
							F05	Delay to Treatment/ Therapy	A0708	Power Problem		
					CT994	Sensors	E060109	Tachycardia	A0708	Power Problem		
							F05	Delay to Treatment/ Therapy	A0708	Power Problem		
13788	2023	All other reportable incidents	47265	Dual-chamber implantable pacemaker, rate-responsive	CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem		
							F1901	Additional Surgery	A24	No Apparent Device Problem		
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem		
							F1901	Additional Surgery	A24	No Apparent Device Problem		
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem		
							F1901	Additional Surgery	A24	No Apparent Device Problem		
13789	2023		17882	Physiologic-monitoring defibrillation system	CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A090204	Image Display Error/Artifact		
							F27	Problem Identified During Non-Clinical Procedure	A090204	Image Display Error/Artifact		
13790	2023	All other reportable incidents	35223	Endocardial/interventricular septal pacing lead	CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070101	Failure to Capture	02/08/2023	02/08/2023
							F1905	Device Revision or Replacement	A070101	Failure to Capture	02/08/2023	02/08/2023
13791	2023	All other reportable incidents	63188	Left-ventricular wireless pacing system sensor/transmitter	CT145	Cardiovascular devices			A0708	Power Problem		
					CT1577	Wireless monitoring system transmitters			A0708	Power Problem		
					CT1684	Cardiac pacemakers and associated devices			A0708	Power Problem		
					CT2727	Active implant devices/ancillaries			A0708	Power Problem		
					CT2750	Implantable cardiac electrical stimulators and associated devices			A0708	Power Problem		
					CT2885	Physiological monitoring/measuring system transmitters/receivers/transceivers			A0708	Power Problem		
13792	2023	Death	60789	Intracardiac pacemaker	CT130	Catheters and associated devices	E2114	Perforation	A24	No Apparent Device Problem	07/08/2023	07/08/2023
							F02	Death	A24	No Apparent Device Problem	07/08/2023	07/08/2023
					CT145	Cardiovascular devices	E2114	Perforation	A24	No Apparent Device Problem	07/08/2023	07/08/2023
							F02	Death	A24	No Apparent Device Problem	07/08/2023	07/08/2023
					CT1684	Cardiac pacemakers and associated devices	E2114	Perforation	A24	No Apparent Device Problem	07/08/2023	07/08/2023
							F02	Death	A24	No Apparent Device Problem	07/08/2023	07/08/2023
					CT1685	Cardiac pacemakers	E2114	Perforation	A24	No Apparent Device Problem	07/08/2023	07/08/2023
							F02	Death	A24	No Apparent Device Problem	07/08/2023	07/08/2023
					CT233	Surgical	E2114	Perforation	A24	No Apparent Device Problem	07/08/2023	07/08/2023
							F02	Death	A24	No Apparent Device Problem	07/08/2023	07/08/2023
					CT2406	Implantable	E2114	Perforation	A24	No Apparent Device Problem	07/08/2023	07/08/2023
							F02	Death	A24	No Apparent Device Problem	07/08/2023	07/08/2023
					CT2407	Active implantable	E2114	Perforation	A24	No Apparent Device Problem	07/08/2023	07/08/2023
							F02	Death	A24	No Apparent Device Problem	07/08/2023	07/08/2023
					CT257	Implantable cardiac pacemakers	E2114	Perforation	A24	No Apparent Device Problem	07/08/2023	07/08/2023

							F02	Death	A24	No Apparent Device Problem	07/08/2023	07/08/2023
					CT2727	Active implant devices/ancillaries	E2114	Perforation	A24	No Apparent Device Problem	07/08/2023	07/08/2023
							F02	Death	A24	No Apparent Device Problem	07/08/2023	07/08/2023
					CT2750	Implantable cardiac electrical stimulators and associated devices	E2114	Perforation	A24	No Apparent Device Problem	07/08/2023	07/08/2023
							F02	Death	A24	No Apparent Device Problem	07/08/2023	07/08/2023
					CT2818	Catheterization/Cannulation/Fluid path devices	E2114	Perforation	A24	No Apparent Device Problem	07/08/2023	07/08/2023
							F02	Death	A24	No Apparent Device Problem	07/08/2023	07/08/2023
					CT2866	Implantable cardiac electrical stimulators	E2114	Perforation	A24	No Apparent Device Problem	07/08/2023	07/08/2023
							F02	Death	A24	No Apparent Device Problem	07/08/2023	07/08/2023
					CT2985	Active	E2114	Perforation	A24	No Apparent Device Problem	07/08/2023	07/08/2023
							F02	Death	A24	No Apparent Device Problem	07/08/2023	07/08/2023
					CT3001	Implantable metal	E2114	Perforation	A24	No Apparent Device Problem	07/08/2023	07/08/2023
							F02	Death	A24	No Apparent Device Problem	07/08/2023	07/08/2023
					CT3002	Implantable synthetic polymer	E2114	Perforation	A24	No Apparent Device Problem	07/08/2023	07/08/2023
							F02	Death	A24	No Apparent Device Problem	07/08/2023	07/08/2023
					CT3006	Implantable inorganic material	E2114	Perforation	A24	No Apparent Device Problem	07/08/2023	07/08/2023
							F02	Death	A24	No Apparent Device Problem	07/08/2023	07/08/2023
					CT314	Electrically powered	E2114	Perforation	A24	No Apparent Device Problem	07/08/2023	07/08/2023
							F02	Death	A24	No Apparent Device Problem	07/08/2023	07/08/2023
					CT321	Long-term surgical invasive	E2114	Perforation	A24	No Apparent Device Problem	07/08/2023	07/08/2023
							F02	Death	A24	No Apparent Device Problem	07/08/2023	07/08/2023
					CT981	Single use	E2114	Perforation	A24	No Apparent Device Problem	07/08/2023	07/08/2023
							F02	Death	A24	No Apparent Device Problem	07/08/2023	07/08/2023
					CT983	Surgical invasive	E2114	Perforation	A24	No Apparent Device Problem	07/08/2023	07/08/2023
							F02	Death	A24	No Apparent Device Problem	07/08/2023	07/08/2023
13793	2023	All other reportable incidents	47265	Dual-chamber implantable pacemaker, rate-responsive	CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A072201	High Impedance	01/10/2022	01/10/2022
							F1901	Additional Surgery	A072201	High Impedance	01/10/2022	01/10/2022
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A072201	High Impedance	01/10/2022	01/10/2022
							F1901	Additional Surgery	A072201	High Impedance	01/10/2022	01/10/2022
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A072201	High Impedance	01/10/2022	01/10/2022
							F1901	Additional Surgery	A072201	High Impedance	01/10/2022	01/10/2022
13794	2023	All other reportable incidents	47205	Cardiac pulse generator programmer	CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A0708	Power Problem		
									A0709	Device Sensing Problem		
									A2204	Inadequate User Interface		
							F23	Unexpected Medical Intervention	A0708	Power Problem		
									A0709	Device Sensing Problem		
									A2204	Inadequate User Interface		
13795	2023	All other reportable incidents	47267	Single-chamber implantable pacemaker, rate-responsive	CT141	Electrodes/Leads and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery		
							F12	Serious Injury/ Illness/ Impairment	A070504	Premature Discharge of Battery		
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery		
							F23	Unexpected Medical Intervention	A070504	Premature Discharge of Battery		
					CT145	Cardiovascular devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery		
							F12	Serious Injury/ Illness/ Impairment	A070504	Premature Discharge of Battery		
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery		
							F23	Unexpected Medical Intervention	A070504	Premature Discharge of Battery		
					CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery		
							F12	Serious Injury/ Illness/ Impairment	A070504	Premature Discharge of Battery		
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery		
							F23	Unexpected Medical Intervention	A070504	Premature Discharge of Battery		
					CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery		
							F12	Serious Injury/ Illness/ Impairment	A070504	Premature Discharge of Battery		
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery		
							F23	Unexpected Medical Intervention	A070504	Premature Discharge of Battery		
					CT233	Surgical	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery		
							F12	Serious Injury/ Illness/ Impairment	A070504	Premature Discharge of Battery		
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery		
							F23	Unexpected Medical Intervention	A070504	Premature Discharge of Battery		
					CT2406	Implantable	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery		
							F12	Serious Injury/ Illness/ Impairment	A070504	Premature Discharge of Battery		
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery		
							F23	Unexpected Medical Intervention	A070504	Premature Discharge of Battery		
					CT2407	Active implantable	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery		
							F12	Serious Injury/ Illness/ Impairment	A070504	Premature Discharge of Battery		
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery		
							F23	Unexpected Medical Intervention	A070504	Premature Discharge of Battery		
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery		
							F12	Serious Injury/ Illness/ Impairment	A070504	Premature Discharge of Battery		
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery		
							F23	Unexpected Medical Intervention	A070504	Premature Discharge of Battery		
					CT2727	Active implant devices/ancillaries	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery		
							F12	Serious Injury/ Illness/ Impairment	A070504	Premature Discharge of Battery		
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery		
							F23	Unexpected Medical Intervention	A070504	Premature Discharge of Battery		
					CT2750	Implantable cardiac electrical stimulators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery		
							F12	Serious Injury/ Illness/ Impairment	A070504	Premature Discharge of Battery		
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery		
							F23	Unexpected Medical Intervention	A070504	Premature Discharge of Battery		
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		

							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery		
							F12	Serious Injury/ Illness/ Impairment	A070504	Premature Discharge of Battery		
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery		
							F23	Unexpected Medical Intervention	A070504	Premature Discharge of Battery		
					CT2985	Active	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery		
							F12	Serious Injury/ Illness/ Impairment	A070504	Premature Discharge of Battery		
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery		
							F23	Unexpected Medical Intervention	A070504	Premature Discharge of Battery		
					CT3001	Implantable metal	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery		
							F12	Serious Injury/ Illness/ Impairment	A070504	Premature Discharge of Battery		
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery		
							F23	Unexpected Medical Intervention	A070504	Premature Discharge of Battery		
					CT3002	Implantable synthetic polymer	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery		
							F12	Serious Injury/ Illness/ Impairment	A070504	Premature Discharge of Battery		
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery		
							F23	Unexpected Medical Intervention	A070504	Premature Discharge of Battery		
					CT3006	Implantable inorganic material	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery		
							F12	Serious Injury/ Illness/ Impairment	A070504	Premature Discharge of Battery		
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery		
							F23	Unexpected Medical Intervention	A070504	Premature Discharge of Battery		
					CT314	Electrically powered	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery		
							F12	Serious Injury/ Illness/ Impairment	A070504	Premature Discharge of Battery		
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery		
							F23	Unexpected Medical Intervention	A070504	Premature Discharge of Battery		
					CT321	Long-term surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery		
							F12	Serious Injury/ Illness/ Impairment	A070504	Premature Discharge of Battery		
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery		
							F23	Unexpected Medical Intervention	A070504	Premature Discharge of Battery		
					CT449	Generators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery		
							F12	Serious Injury/ Illness/ Impairment	A070504	Premature Discharge of Battery		
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery		
							F23	Unexpected Medical Intervention	A070504	Premature Discharge of Battery		
					CT981	Single use	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery		
							F12	Serious Injury/ Illness/ Impairment	A070504	Premature Discharge of Battery		
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery		
							F23	Unexpected Medical Intervention	A070504	Premature Discharge of Battery		
					CT983	Surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery		
							F12	Serious Injury/ Illness/ Impairment	A070504	Premature Discharge of Battery		
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery		
							F23	Unexpected Medical Intervention	A070504	Premature Discharge of Battery		
13796	2023	All other reportable incidents	47265	Dual-chamber implantable pacemaker, rate-responsive	CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F1901	Additional Surgery	A070504	Premature Discharge of Battery		
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F1901	Additional Surgery	A070504	Premature Discharge of Battery		
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F1901	Additional Surgery	A070504	Premature Discharge of Battery		
13797	2023	All other reportable incidents	47270	Cardiac resynchronization therapy implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
13798	2023	All other reportable incidents	47267	Single-chamber implantable pacemaker, rate-responsive	CT141	Electrodes/Leads and associated devices	E0601	Arrhythmia	A071102	Failure to Interrogate	10/08/2023	10/08/2023
									A071204	No Pacing	10/08/2023	10/08/2023
							F1901	Additional Surgery	A071102	Failure to Interrogate	10/08/2023	10/08/2023
									A071204	No Pacing	10/08/2023	10/08/2023
					CT145	Cardiovascular devices	E0601	Arrhythmia	A071102	Failure to Interrogate	10/08/2023	10/08/2023
									A071204	No Pacing	10/08/2023	10/08/2023
							F1901	Additional Surgery	A071102	Failure to Interrogate	10/08/2023	10/08/2023
									A071204	No Pacing	10/08/2023	10/08/2023
					CT1684	Cardiac pacemakers and associated devices	E0601	Arrhythmia	A071102	Failure to Interrogate	10/08/2023	10/08/2023
									A071204	No Pacing	10/08/2023	10/08/2023
							F1901	Additional Surgery	A071102	Failure to Interrogate	10/08/2023	10/08/2023
									A071204	No Pacing	10/08/2023	10/08/2023
					CT1685	Cardiac pacemakers	E0601	Arrhythmia	A071102	Failure to Interrogate	10/08/2023	10/08/2023
									A071204	No Pacing	10/08/2023	10/08/2023
							F1901	Additional Surgery	A071102	Failure to Interrogate	10/08/2023	10/08/2023
									A071204	No Pacing	10/08/2023	10/08/2023
					CT233	Surgical	E0601	Arrhythmia	A071102	Failure to Interrogate	10/08/2023	10/08/2023
									A071204	No Pacing	10/08/2023	10/08/2023
							F1901	Additional Surgery	A071102	Failure to Interrogate	10/08/2023	10/08/2023
									A071204	No Pacing	10/08/2023	10/08/2023
					CT2406	Implantable	E0601	Arrhythmia	A071102	Failure to Interrogate	10/08/2023	10/08/2023
									A071204	No Pacing	10/08/2023	10/08/2023
							F1901	Additional Surgery	A071102	Failure to Interrogate	10/08/2023	10/08/2023
									A071204	No Pacing	10/08/2023	10/08/2023
					CT2407	Active implantable	E0601	Arrhythmia	A071102	Failure to Interrogate	10/08/2023	10/08/2023
									A071204	No Pacing	10/08/2023	10/08/2023
							F1901	Additional Surgery	A071102	Failure to Interrogate	10/08/2023	10/08/2023
									A071204	No Pacing	10/08/2023	10/08/2023
					CT257	Implantable cardiac pacemakers	E0601	Arrhythmia	A071102	Failure to Interrogate	10/08/2023	10/08/2023
									A071204	No Pacing	10/08/2023	10/08/2023
							F1901	Additional Surgery	A071102	Failure to Interrogate	10/08/2023	10/08/2023

									A071204	No Pacing	10/08/2023	10/08/2023
					CT2727	Active implant devices/ancillaries	E0601	Arrhythmia	A071102	Failure to Interrogate	10/08/2023	10/08/2023
									A071204	No Pacing	10/08/2023	10/08/2023
							F1901	Additional Surgery	A071102	Failure to Interrogate	10/08/2023	10/08/2023
									A071204	No Pacing	10/08/2023	10/08/2023
					CT2750	Implantable cardiac electrical stimulators and associated devices	E0601	Arrhythmia	A071102	Failure to Interrogate	10/08/2023	10/08/2023
									A071204	No Pacing	10/08/2023	10/08/2023
							F1901	Additional Surgery	A071102	Failure to Interrogate	10/08/2023	10/08/2023
									A071204	No Pacing	10/08/2023	10/08/2023
					CT2866	Implantable cardiac electrical stimulators	E0601	Arrhythmia	A071102	Failure to Interrogate	10/08/2023	10/08/2023
									A071204	No Pacing	10/08/2023	10/08/2023
							F1901	Additional Surgery	A071102	Failure to Interrogate	10/08/2023	10/08/2023
									A071204	No Pacing	10/08/2023	10/08/2023
					CT2985	Active	E0601	Arrhythmia	A071102	Failure to Interrogate	10/08/2023	10/08/2023
									A071204	No Pacing	10/08/2023	10/08/2023
							F1901	Additional Surgery	A071102	Failure to Interrogate	10/08/2023	10/08/2023
									A071204	No Pacing	10/08/2023	10/08/2023
					CT3001	Implantable metal	E0601	Arrhythmia	A071102	Failure to Interrogate	10/08/2023	10/08/2023
									A071204	No Pacing	10/08/2023	10/08/2023
							F1901	Additional Surgery	A071102	Failure to Interrogate	10/08/2023	10/08/2023
									A071204	No Pacing	10/08/2023	10/08/2023
					CT3002	Implantable synthetic polymer	E0601	Arrhythmia	A071102	Failure to Interrogate	10/08/2023	10/08/2023
									A071204	No Pacing	10/08/2023	10/08/2023
							F1901	Additional Surgery	A071102	Failure to Interrogate	10/08/2023	10/08/2023
									A071204	No Pacing	10/08/2023	10/08/2023
					CT3006	Implantable inorganic material	E0601	Arrhythmia	A071102	Failure to Interrogate	10/08/2023	10/08/2023
									A071204	No Pacing	10/08/2023	10/08/2023
							F1901	Additional Surgery	A071102	Failure to Interrogate	10/08/2023	10/08/2023
									A071204	No Pacing	10/08/2023	10/08/2023
					CT314	Electrically powered	E0601	Arrhythmia	A071102	Failure to Interrogate	10/08/2023	10/08/2023
									A071204	No Pacing	10/08/2023	10/08/2023
							F1901	Additional Surgery	A071102	Failure to Interrogate	10/08/2023	10/08/2023
									A071204	No Pacing	10/08/2023	10/08/2023
					CT321	Long-term surgical invasive	E0601	Arrhythmia	A071102	Failure to Interrogate	10/08/2023	10/08/2023
									A071204	No Pacing	10/08/2023	10/08/2023
							F1901	Additional Surgery	A071102	Failure to Interrogate	10/08/2023	10/08/2023
									A071204	No Pacing	10/08/2023	10/08/2023
					CT449	Generators and associated devices	E0601	Arrhythmia	A071102	Failure to Interrogate	10/08/2023	10/08/2023
									A071204	No Pacing	10/08/2023	10/08/2023
							F1901	Additional Surgery	A071102	Failure to Interrogate	10/08/2023	10/08/2023
									A071204	No Pacing	10/08/2023	10/08/2023
					CT981	Single use	E0601	Arrhythmia	A071102	Failure to Interrogate	10/08/2023	10/08/2023
									A071204	No Pacing	10/08/2023	10/08/2023
							F1901	Additional Surgery	A071102	Failure to Interrogate	10/08/2023	10/08/2023
									A071204	No Pacing	10/08/2023	10/08/2023
					CT983	Surgical invasive	E0601	Arrhythmia	A071102	Failure to Interrogate	10/08/2023	10/08/2023
									A071204	No Pacing	10/08/2023	10/08/2023
							F1901	Additional Surgery	A071102	Failure to Interrogate	10/08/2023	10/08/2023
									A071204	No Pacing	10/08/2023	10/08/2023
13799	2023	All other reportable incidents	47265	Dual-chamber implantable pacemaker, rate-responsive	CT141	Electrodes/Leads and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070909	Oversensing	03/08/2023	03/08/2023
									A0723	Unexpected Mode Switch	03/08/2023	03/08/2023
							F1901	Additional Surgery	A070909	Oversensing	03/08/2023	03/08/2023
									A0723	Unexpected Mode Switch	03/08/2023	03/08/2023
					CT145	Cardiovascular devices	E2403	No Clinical Signs, Symptoms or Conditions	A070909	Oversensing	03/08/2023	03/08/2023
									A0723	Unexpected Mode Switch	03/08/2023	03/08/2023
							F1901	Additional Surgery	A070909	Oversensing	03/08/2023	03/08/2023
									A0723	Unexpected Mode Switch	03/08/2023	03/08/2023
					CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070909	Oversensing	03/08/2023	03/08/2023
									A0723	Unexpected Mode Switch	03/08/2023	03/08/2023
							F1901	Additional Surgery	A070909	Oversensing	03/08/2023	03/08/2023
									A0723	Unexpected Mode Switch	03/08/2023	03/08/2023
					CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A070909	Oversensing	03/08/2023	03/08/2023
									A0723	Unexpected Mode Switch	03/08/2023	03/08/2023
							F1901	Additional Surgery	A070909	Oversensing	03/08/2023	03/08/2023
									A0723	Unexpected Mode Switch	03/08/2023	03/08/2023
					CT233	Surgical	E2403	No Clinical Signs, Symptoms or Conditions	A070909	Oversensing	03/08/2023	03/08/2023
									A0723	Unexpected Mode Switch	03/08/2023	03/08/2023
							F1901	Additional Surgery	A070909	Oversensing	03/08/2023	03/08/2023
									A0723	Unexpected Mode Switch	03/08/2023	03/08/2023
					CT2406	Implantable	E2403	No Clinical Signs, Symptoms or Conditions	A070909	Oversensing	03/08/2023	03/08/2023
									A0723	Unexpected Mode Switch	03/08/2023	03/08/2023
							F1901	Additional Surgery	A070909	Oversensing	03/08/2023	03/08/2023
									A0723	Unexpected Mode Switch	03/08/2023	03/08/2023
					CT2407	Active implantable	E2403	No Clinical Signs, Symptoms or Conditions	A070909	Oversensing	03/08/2023	03/08/2023
									A0723	Unexpected Mode Switch	03/08/2023	03/08/2023
							F1901	Additional Surgery	A070909	Oversensing	03/08/2023	03/08/2023
									A0723	Unexpected Mode Switch	03/08/2023	03/08/2023
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A070909	Oversensing	03/08/2023	03/08/2023
									A0723	Unexpected Mode Switch	03/08/2023	03/08/2023
							F1901	Additional Surgery	A070909	Oversensing	03/08/2023	03/08/2023
									A0723	Unexpected Mode Switch	03/08/2023	03/08/2023
					CT2727	Active implant devices/ancillaries	E2403	No Clinical Signs, Symptoms or Conditions	A070909	Oversensing	03/08/2023	03/08/2023
									A0723	Unexpected Mode Switch	03/08/2023	03/08/2023
							F1901	Additional Surgery	A070909	Oversensing	03/08/2023	03/08/2023
									A0723	Unexpected Mode Switch	03/08/2023	03/08/2023
					CT2750	Implantable cardiac electrical stimulators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070909	Oversensing	03/08/2023	03/08/2023
									A0723	Unexpected Mode Switch	03/08/2023	03/08/2023
							F1901	Additional Surgery	A070909	Oversensing	03/08/2023	03/08/2023

									A0723	Unexpected Mode Switch	03/08/2023	03/08/2023
				CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A070909	Oversensing	03/08/2023	03/08/2023
								A0723	Unexpected Mode Switch	03/08/2023	03/08/2023
						F1901	Additional Surgery	A070909	Oversensing	03/08/2023	03/08/2023
								A0723	Unexpected Mode Switch	03/08/2023	03/08/2023
				CT2985	Active	E2403	No Clinical Signs, Symptoms or Conditions	A070909	Oversensing	03/08/2023	03/08/2023
								A0723	Unexpected Mode Switch	03/08/2023	03/08/2023
						F1901	Additional Surgery	A070909	Oversensing	03/08/2023	03/08/2023
								A0723	Unexpected Mode Switch	03/08/2023	03/08/2023
				CT3001	Implantable metal	E2403	No Clinical Signs, Symptoms or Conditions	A070909	Oversensing	03/08/2023	03/08/2023
								A0723	Unexpected Mode Switch	03/08/2023	03/08/2023
						F1901	Additional Surgery	A070909	Oversensing	03/08/2023	03/08/2023
								A0723	Unexpected Mode Switch	03/08/2023	03/08/2023
				CT3002	Implantable synthetic polymer	E2403	No Clinical Signs, Symptoms or Conditions	A070909	Oversensing	03/08/2023	03/08/2023
								A0723	Unexpected Mode Switch	03/08/2023	03/08/2023
						F1901	Additional Surgery	A070909	Oversensing	03/08/2023	03/08/2023
								A0723	Unexpected Mode Switch	03/08/2023	03/08/2023
				CT3006	Implantable inorganic material	E2403	No Clinical Signs, Symptoms or Conditions	A070909	Oversensing	03/08/2023	03/08/2023
								A0723	Unexpected Mode Switch	03/08/2023	03/08/2023
						F1901	Additional Surgery	A070909	Oversensing	03/08/2023	03/08/2023
								A0723	Unexpected Mode Switch	03/08/2023	03/08/2023
				CT314	Electrically powered	E2403	No Clinical Signs, Symptoms or Conditions	A070909	Oversensing	03/08/2023	03/08/2023
								A0723	Unexpected Mode Switch	03/08/2023	03/08/2023
						F1901	Additional Surgery	A070909	Oversensing	03/08/2023	03/08/2023
								A0723	Unexpected Mode Switch	03/08/2023	03/08/2023
				CT321	Long-term surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A070909	Oversensing	03/08/2023	03/08/2023
								A0723	Unexpected Mode Switch	03/08/2023	03/08/2023
						F1901	Additional Surgery	A070909	Oversensing	03/08/2023	03/08/2023
								A0723	Unexpected Mode Switch	03/08/2023	03/08/2023
				CT449	Generators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070909	Oversensing	03/08/2023	03/08/2023
								A0723	Unexpected Mode Switch	03/08/2023	03/08/2023
						F1901	Additional Surgery	A070909	Oversensing	03/08/2023	03/08/2023
								A0723	Unexpected Mode Switch	03/08/2023	03/08/2023
				CT981	Single use	E2403	No Clinical Signs, Symptoms or Conditions	A070909	Oversensing	03/08/2023	03/08/2023
								A0723	Unexpected Mode Switch	03/08/2023	03/08/2023
						F1901	Additional Surgery	A070909	Oversensing	03/08/2023	03/08/2023
								A0723	Unexpected Mode Switch	03/08/2023	03/08/2023
				CT983	Surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A070909	Oversensing	03/08/2023	03/08/2023
								A0723	Unexpected Mode Switch	03/08/2023	03/08/2023
						F1901	Additional Surgery	A070909	Oversensing	03/08/2023	03/08/2023
								A0723	Unexpected Mode Switch	03/08/2023	03/08/2023
13800	2023	All other reportable incidents	47265	Dual-chamber implantable pacemaker, rate-responsive	CT141	Electrodes/Leads and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070909	Oversensing		
							F26	No Health Consequences or Impact	A070909	Oversensing		
					CT145	Cardiovascular devices	E2403	No Clinical Signs, Symptoms or Conditions	A070909	Oversensing		
							F26	No Health Consequences or Impact	A070909	Oversensing		
					CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070909	Oversensing		
							F26	No Health Consequences or Impact	A070909	Oversensing		
					CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A070909	Oversensing		
							F26	No Health Consequences or Impact	A070909	Oversensing		
					CT233	Surgical	E2403	No Clinical Signs, Symptoms or Conditions	A070909	Oversensing		
							F26	No Health Consequences or Impact	A070909	Oversensing		
					CT2406	Implantable	E2403	No Clinical Signs, Symptoms or Conditions	A070909	Oversensing		
							F26	No Health Consequences or Impact	A070909	Oversensing		
					CT2407	Active implantable	E2403	No Clinical Signs, Symptoms or Conditions	A070909	Oversensing		
							F26	No Health Consequences or Impact	A070909	Oversensing		
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A070909	Oversensing		
							F26	No Health Consequences or Impact	A070909	Oversensing		
					CT2727	Active implant devices/ancillaries	E2403	No Clinical Signs, Symptoms or Conditions	A070909	Oversensing		
							F26	No Health Consequences or Impact	A070909	Oversensing		
					CT2750	Implantable cardiac electrical stimulators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070909	Oversensing		
							F26	No Health Consequences or Impact	A070909	Oversensing		
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A070909	Oversensing		
							F26	No Health Consequences or Impact	A070909	Oversensing		
					CT2985	Active	E2403	No Clinical Signs, Symptoms or Conditions	A070909	Oversensing		
							F26	No Health Consequences or Impact	A070909	Oversensing		
					CT3001	Implantable metal	E2403	No Clinical Signs, Symptoms or Conditions	A070909	Oversensing		
							F26	No Health Consequences or Impact	A070909	Oversensing		
					CT3002	Implantable synthetic polymer	E2403	No Clinical Signs, Symptoms or Conditions	A070909	Oversensing		
							F26	No Health Consequences or Impact	A070909	Oversensing		
					CT3006	Implantable inorganic material	E2403	No Clinical Signs, Symptoms or Conditions	A070909	Oversensing		
							F26	No Health Consequences or Impact	A070909	Oversensing		
					CT314	Electrically powered	E2403	No Clinical Signs, Symptoms or Conditions	A070909	Oversensing		
							F26	No Health Consequences or Impact	A070909	Oversensing		
					CT321	Long-term surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A070909	Oversensing		
							F26	No Health Consequences or Impact	A070909	Oversensing		
					CT449	Generators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070909	Oversensing		
							F26	No Health Consequences or Impact	A070909	Oversensing		
					CT981	Single use	E2403	No Clinical Signs, Symptoms or Conditions	A070909	Oversensing		
							F26	No Health Consequences or Impact	A070909	Oversensing		
					CT983	Surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A070909	Oversensing		
							F26	No Health Consequences or Impact	A070909	Oversensing		
13801	2023	All other reportable incidents	47267	Single-chamber implantable pacemaker, rate-responsive	CT141	Electrodes/Leads and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F27	Problem Identified During Non-Clinical Procedure	A070504	Premature Discharge of Battery		
					CT145	Cardiovascular devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F27	Problem Identified During Non-Clinical Procedure	A070504	Premature Discharge of Battery		
					CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F27	Problem Identified During Non-Clinical Procedure	A070504	Premature Discharge of Battery		
					CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		

							F27	Problem Identified During Non-Clinical Procedure	A070504	Premature Discharge of Battery		
					CT233	Surgical	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F27	Problem Identified During Non-Clinical Procedure	A070504	Premature Discharge of Battery		
					CT2406	Implantable	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F27	Problem Identified During Non-Clinical Procedure	A070504	Premature Discharge of Battery		
					CT2407	Active implantable	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F27	Problem Identified During Non-Clinical Procedure	A070504	Premature Discharge of Battery		
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F27	Problem Identified During Non-Clinical Procedure	A070504	Premature Discharge of Battery		
					CT2727	Active implant devices/ancillaries	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F27	Problem Identified During Non-Clinical Procedure	A070504	Premature Discharge of Battery		
					CT2750	Implantable cardiac electrical stimulators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F27	Problem Identified During Non-Clinical Procedure	A070504	Premature Discharge of Battery		
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F27	Problem Identified During Non-Clinical Procedure	A070504	Premature Discharge of Battery		
					CT2985	Active	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F27	Problem Identified During Non-Clinical Procedure	A070504	Premature Discharge of Battery		
					CT3001	Implantable metal	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F27	Problem Identified During Non-Clinical Procedure	A070504	Premature Discharge of Battery		
					CT3002	Implantable synthetic polymer	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F27	Problem Identified During Non-Clinical Procedure	A070504	Premature Discharge of Battery		
					CT3006	Implantable inorganic material	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F27	Problem Identified During Non-Clinical Procedure	A070504	Premature Discharge of Battery		
					CT314	Electrically powered	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F27	Problem Identified During Non-Clinical Procedure	A070504	Premature Discharge of Battery		
					CT321	Long-term surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F27	Problem Identified During Non-Clinical Procedure	A070504	Premature Discharge of Battery		
					CT449	Generators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F27	Problem Identified During Non-Clinical Procedure	A070504	Premature Discharge of Battery		
					CT981	Single use	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F27	Problem Identified During Non-Clinical Procedure	A070504	Premature Discharge of Battery		
					CT983	Surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F27	Problem Identified During Non-Clinical Procedure	A070504	Premature Discharge of Battery		
13802	2023	All other reportable incidents	35852	Single-chamber implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	28/03/2023	28/03/2023
									A0908	Incorrect, Inadequate or Imprecise Result or Readings	28/03/2023	28/03/2023
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	28/03/2023	28/03/2023
									A0908	Incorrect, Inadequate or Imprecise Result or Readings	28/03/2023	28/03/2023
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	28/03/2023	28/03/2023
									A0908	Incorrect, Inadequate or Imprecise Result or Readings	28/03/2023	28/03/2023
13803	2023	All other reportable incidents	47267	Single-chamber implantable pacemaker, rate-responsive	CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem	18/10/2023	18/10/2023
									A070504	Premature Discharge of Battery	18/10/2023	18/10/2023
							F08	Hospitalization or Prolonged Hospitalization	A0705	Battery Problem	18/10/2023	18/10/2023
									A070504	Premature Discharge of Battery	18/10/2023	18/10/2023
							F1903	Device Explantation	A0705	Battery Problem	18/10/2023	18/10/2023
									A070504	Premature Discharge of Battery	18/10/2023	18/10/2023
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem	18/10/2023	18/10/2023
									A070504	Premature Discharge of Battery	18/10/2023	18/10/2023
							F08	Hospitalization or Prolonged Hospitalization	A0705	Battery Problem	18/10/2023	18/10/2023
									A070504	Premature Discharge of Battery	18/10/2023	18/10/2023
							F1903	Device Explantation	A0705	Battery Problem	18/10/2023	18/10/2023
									A070504	Premature Discharge of Battery	18/10/2023	18/10/2023
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem	18/10/2023	18/10/2023
									A070504	Premature Discharge of Battery	18/10/2023	18/10/2023
							F08	Hospitalization or Prolonged Hospitalization	A0705	Battery Problem	18/10/2023	18/10/2023
									A070504	Premature Discharge of Battery	18/10/2023	18/10/2023
							F1903	Device Explantation	A0705	Battery Problem	18/10/2023	18/10/2023
									A070504	Premature Discharge of Battery	18/10/2023	18/10/2023
13804	2023	All other reportable incidents	47265	Dual-chamber implantable pacemaker, rate-responsive	CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A070102	High Capture Threshold	01/06/2023	01/06/2023
									A0705	Battery Problem	01/06/2023	01/06/2023
									A070504	Premature Discharge of Battery	01/06/2023	01/06/2023
									A070910	Undersensing	01/06/2023	01/06/2023
									A0712	Pacing Problem	01/06/2023	01/06/2023
									A072201	High Impedance	01/06/2023	01/06/2023
							F08	Hospitalization or Prolonged Hospitalization	A070102	High Capture Threshold	01/06/2023	01/06/2023
									A0705	Battery Problem	01/06/2023	01/06/2023
									A070504	Premature Discharge of Battery	01/06/2023	01/06/2023
									A070910	Undersensing	01/06/2023	01/06/2023
									A0712	Pacing Problem	01/06/2023	01/06/2023
									A072201	High Impedance	01/06/2023	01/06/2023
							F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A070102	High Capture Threshold	01/06/2023	01/06/2023
									A0705	Battery Problem	01/06/2023	01/06/2023
									A070504	Premature Discharge of Battery	01/06/2023	01/06/2023
									A070910	Undersensing	01/06/2023	01/06/2023
									A0712	Pacing Problem	01/06/2023	01/06/2023
									A072201	High Impedance	01/06/2023	01/06/2023
							F1905	Device Revision or Replacement	A070102	High Capture Threshold	01/06/2023	01/06/2023
									A0705	Battery Problem	01/06/2023	01/06/2023
									A070504	Premature Discharge of Battery	01/06/2023	01/06/2023
									A070910	Undersensing	01/06/2023	01/06/2023
									A0712	Pacing Problem	01/06/2023	01/06/2023
									A072201	High Impedance	01/06/2023	01/06/2023
							F2203	Imaging Required	A070102	High Capture Threshold	01/06/2023	01/06/2023
									A0705	Battery Problem	01/06/2023	01/06/2023
									A070504	Premature Discharge of Battery	01/06/2023	01/06/2023
									A070910	Undersensing	01/06/2023	01/06/2023

								A0712	Pacing Problem	01/06/2023	01/06/2023
								A072201	High Impedance	01/06/2023	01/06/2023
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A070102	High Capture Threshold	01/06/2023	01/06/2023
									A0705	Battery Problem	01/06/2023	01/06/2023
									A070504	Premature Discharge of Battery	01/06/2023	01/06/2023
									A070910	Undersensing	01/06/2023	01/06/2023
									A0712	Pacing Problem	01/06/2023	01/06/2023
									A072201	High Impedance	01/06/2023	01/06/2023
							F08	Hospitalization or Prolonged Hospitalization	A070102	High Capture Threshold	01/06/2023	01/06/2023
									A0705	Battery Problem	01/06/2023	01/06/2023
									A070504	Premature Discharge of Battery	01/06/2023	01/06/2023
									A070910	Undersensing	01/06/2023	01/06/2023
									A0712	Pacing Problem	01/06/2023	01/06/2023
									A072201	High Impedance	01/06/2023	01/06/2023
							F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A070102	High Capture Threshold	01/06/2023	01/06/2023
									A0705	Battery Problem	01/06/2023	01/06/2023
									A070504	Premature Discharge of Battery	01/06/2023	01/06/2023
									A070910	Undersensing	01/06/2023	01/06/2023
									A0712	Pacing Problem	01/06/2023	01/06/2023
									A072201	High Impedance	01/06/2023	01/06/2023
							F1905	Device Revision or Replacement	A070102	High Capture Threshold	01/06/2023	01/06/2023
									A0705	Battery Problem	01/06/2023	01/06/2023
									A070504	Premature Discharge of Battery	01/06/2023	01/06/2023
									A070910	Undersensing	01/06/2023	01/06/2023
									A0712	Pacing Problem	01/06/2023	01/06/2023
									A072201	High Impedance	01/06/2023	01/06/2023
							F2203	Imaging Required	A070102	High Capture Threshold	01/06/2023	01/06/2023
									A0705	Battery Problem	01/06/2023	01/06/2023
									A070504	Premature Discharge of Battery	01/06/2023	01/06/2023
									A070910	Undersensing	01/06/2023	01/06/2023
									A0712	Pacing Problem	01/06/2023	01/06/2023
									A072201	High Impedance	01/06/2023	01/06/2023
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A070102	High Capture Threshold	01/06/2023	01/06/2023
									A0705	Battery Problem	01/06/2023	01/06/2023
									A070504	Premature Discharge of Battery	01/06/2023	01/06/2023
									A070910	Undersensing	01/06/2023	01/06/2023
									A0712	Pacing Problem	01/06/2023	01/06/2023
									A072201	High Impedance	01/06/2023	01/06/2023
							F08	Hospitalization or Prolonged Hospitalization	A070102	High Capture Threshold	01/06/2023	01/06/2023
									A0705	Battery Problem	01/06/2023	01/06/2023
									A070504	Premature Discharge of Battery	01/06/2023	01/06/2023
									A070910	Undersensing	01/06/2023	01/06/2023
									A0712	Pacing Problem	01/06/2023	01/06/2023
									A072201	High Impedance	01/06/2023	01/06/2023
							F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A070102	High Capture Threshold	01/06/2023	01/06/2023
									A0705	Battery Problem	01/06/2023	01/06/2023
									A070504	Premature Discharge of Battery	01/06/2023	01/06/2023
									A070910	Undersensing	01/06/2023	01/06/2023
									A0712	Pacing Problem	01/06/2023	01/06/2023
									A072201	High Impedance	01/06/2023	01/06/2023
							F1905	Device Revision or Replacement	A070102	High Capture Threshold	01/06/2023	01/06/2023
									A0705	Battery Problem	01/06/2023	01/06/2023
									A070504	Premature Discharge of Battery	01/06/2023	01/06/2023
									A070910	Undersensing	01/06/2023	01/06/2023
									A0712	Pacing Problem	01/06/2023	01/06/2023
									A072201	High Impedance	01/06/2023	01/06/2023
							F2203	Imaging Required	A070102	High Capture Threshold	01/06/2023	01/06/2023
									A0705	Battery Problem	01/06/2023	01/06/2023
									A070504	Premature Discharge of Battery	01/06/2023	01/06/2023
									A070910	Undersensing	01/06/2023	01/06/2023
									A0712	Pacing Problem	01/06/2023	01/06/2023
									A072201	High Impedance	01/06/2023	01/06/2023
13805	2023	All other reportable incidents	47263	Cardiac resynchronization therapy implantable pacemaker	CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A0908	Incorrect, Inadequate or Imprecise Result or Readings	03/08/2023	03/08/2023
							F08	Hospitalization or Prolonged Hospitalization	A0908	Incorrect, Inadequate or Imprecise Result or Readings	03/08/2023	03/08/2023
							F1905	Device Revision or Replacement	A0908	Incorrect, Inadequate or Imprecise Result or Readings	03/08/2023	03/08/2023
					CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A0908	Incorrect, Inadequate or Imprecise Result or Readings	03/08/2023	03/08/2023
							F08	Hospitalization or Prolonged Hospitalization	A0908	Incorrect, Inadequate or Imprecise Result or Readings	03/08/2023	03/08/2023
							F1905	Device Revision or Replacement	A0908	Incorrect, Inadequate or Imprecise Result or Readings	03/08/2023	03/08/2023
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A0908	Incorrect, Inadequate or Imprecise Result or Readings	03/08/2023	03/08/2023
							F08	Hospitalization or Prolonged Hospitalization	A0908	Incorrect, Inadequate or Imprecise Result or Readings	03/08/2023	03/08/2023
							F1905	Device Revision or Replacement	A0908	Incorrect, Inadequate or Imprecise Result or Readings	03/08/2023	03/08/2023
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A0908	Incorrect, Inadequate or Imprecise Result or Readings	03/08/2023	03/08/2023
							F08	Hospitalization or Prolonged Hospitalization	A0908	Incorrect, Inadequate or Imprecise Result or Readings	03/08/2023	03/08/2023
							F1905	Device Revision or Replacement	A0908	Incorrect, Inadequate or Imprecise Result or Readings	03/08/2023	03/08/2023
13806	2023	All other reportable incidents	47265	Dual-chamber implantable pacemaker, rate-responsive	CT141	Electrodes/Leads and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A0908	Incorrect, Inadequate or Imprecise Result or Readings	12/05/2023	12/05/2023
							F08	Hospitalization or Prolonged Hospitalization	A0908	Incorrect, Inadequate or Imprecise Result or Readings	12/05/2023	12/05/2023

						F1905	Device Revision or Replacement	A0908	Incorrect, Inadequate or Imprecise Result or Readings	12/05/2023	12/05/2023
					CT145	E2403	No Clinical Signs, Symptoms or Conditions	A0908	Incorrect, Inadequate or Imprecise Result or Readings	12/05/2023	12/05/2023
						F08	Hospitalization or Prolonged Hospitalization	A0908	Incorrect, Inadequate or Imprecise Result or Readings	12/05/2023	12/05/2023
						F1905	Device Revision or Replacement	A0908	Incorrect, Inadequate or Imprecise Result or Readings	12/05/2023	12/05/2023
					CT1684	E2403	No Clinical Signs, Symptoms or Conditions	A0908	Incorrect, Inadequate or Imprecise Result or Readings	12/05/2023	12/05/2023
						F08	Hospitalization or Prolonged Hospitalization	A0908	Incorrect, Inadequate or Imprecise Result or Readings	12/05/2023	12/05/2023
						F1905	Device Revision or Replacement	A0908	Incorrect, Inadequate or Imprecise Result or Readings	12/05/2023	12/05/2023
					CT1685	E2403	No Clinical Signs, Symptoms or Conditions	A0908	Incorrect, Inadequate or Imprecise Result or Readings	12/05/2023	12/05/2023
						F08	Hospitalization or Prolonged Hospitalization	A0908	Incorrect, Inadequate or Imprecise Result or Readings	12/05/2023	12/05/2023
						F1905	Device Revision or Replacement	A0908	Incorrect, Inadequate or Imprecise Result or Readings	12/05/2023	12/05/2023
					CT233	E2403	No Clinical Signs, Symptoms or Conditions	A0908	Incorrect, Inadequate or Imprecise Result or Readings	12/05/2023	12/05/2023
						F08	Hospitalization or Prolonged Hospitalization	A0908	Incorrect, Inadequate or Imprecise Result or Readings	12/05/2023	12/05/2023
						F1905	Device Revision or Replacement	A0908	Incorrect, Inadequate or Imprecise Result or Readings	12/05/2023	12/05/2023
					CT2406	E2403	No Clinical Signs, Symptoms or Conditions	A0908	Incorrect, Inadequate or Imprecise Result or Readings	12/05/2023	12/05/2023
						F08	Hospitalization or Prolonged Hospitalization	A0908	Incorrect, Inadequate or Imprecise Result or Readings	12/05/2023	12/05/2023
						F1905	Device Revision or Replacement	A0908	Incorrect, Inadequate or Imprecise Result or Readings	12/05/2023	12/05/2023
					CT2407	E2403	No Clinical Signs, Symptoms or Conditions	A0908	Incorrect, Inadequate or Imprecise Result or Readings	12/05/2023	12/05/2023
						F08	Hospitalization or Prolonged Hospitalization	A0908	Incorrect, Inadequate or Imprecise Result or Readings	12/05/2023	12/05/2023
						F1905	Device Revision or Replacement	A0908	Incorrect, Inadequate or Imprecise Result or Readings	12/05/2023	12/05/2023
					CT257	E2403	No Clinical Signs, Symptoms or Conditions	A0908	Incorrect, Inadequate or Imprecise Result or Readings	12/05/2023	12/05/2023
						F08	Hospitalization or Prolonged Hospitalization	A0908	Incorrect, Inadequate or Imprecise Result or Readings	12/05/2023	12/05/2023
						F1905	Device Revision or Replacement	A0908	Incorrect, Inadequate or Imprecise Result or Readings	12/05/2023	12/05/2023
					CT2727	E2403	No Clinical Signs, Symptoms or Conditions	A0908	Incorrect, Inadequate or Imprecise Result or Readings	12/05/2023	12/05/2023
						F08	Hospitalization or Prolonged Hospitalization	A0908	Incorrect, Inadequate or Imprecise Result or Readings	12/05/2023	12/05/2023
						F1905	Device Revision or Replacement	A0908	Incorrect, Inadequate or Imprecise Result or Readings	12/05/2023	12/05/2023
					CT2750	E2403	No Clinical Signs, Symptoms or Conditions	A0908	Incorrect, Inadequate or Imprecise Result or Readings	12/05/2023	12/05/2023
						F08	Hospitalization or Prolonged Hospitalization	A0908	Incorrect, Inadequate or Imprecise Result or Readings	12/05/2023	12/05/2023
						F1905	Device Revision or Replacement	A0908	Incorrect, Inadequate or Imprecise Result or Readings	12/05/2023	12/05/2023
					CT2866	E2403	No Clinical Signs, Symptoms or Conditions	A0908	Incorrect, Inadequate or Imprecise Result or Readings	12/05/2023	12/05/2023
						F08	Hospitalization or Prolonged Hospitalization	A0908	Incorrect, Inadequate or Imprecise Result or Readings	12/05/2023	12/05/2023
						F1905	Device Revision or Replacement	A0908	Incorrect, Inadequate or Imprecise Result or Readings	12/05/2023	12/05/2023
					CT2985	E2403	No Clinical Signs, Symptoms or Conditions	A0908	Incorrect, Inadequate or Imprecise Result or Readings	12/05/2023	12/05/2023
						F08	Hospitalization or Prolonged Hospitalization	A0908	Incorrect, Inadequate or Imprecise Result or Readings	12/05/2023	12/05/2023
						F1905	Device Revision or Replacement	A0908	Incorrect, Inadequate or Imprecise Result or Readings	12/05/2023	12/05/2023
					CT3001	E2403	No Clinical Signs, Symptoms or Conditions	A0908	Incorrect, Inadequate or Imprecise Result or Readings	12/05/2023	12/05/2023
						F08	Hospitalization or Prolonged Hospitalization	A0908	Incorrect, Inadequate or Imprecise Result or Readings	12/05/2023	12/05/2023
						F1905	Device Revision or Replacement	A0908	Incorrect, Inadequate or Imprecise Result or Readings	12/05/2023	12/05/2023
					CT3002	E2403	No Clinical Signs, Symptoms or Conditions	A0908	Incorrect, Inadequate or Imprecise Result or Readings	12/05/2023	12/05/2023
						F08	Hospitalization or Prolonged Hospitalization	A0908	Incorrect, Inadequate or Imprecise Result or Readings	12/05/2023	12/05/2023
						F1905	Device Revision or Replacement	A0908	Incorrect, Inadequate or Imprecise Result or Readings	12/05/2023	12/05/2023
					CT3006	E2403	No Clinical Signs, Symptoms or Conditions	A0908	Incorrect, Inadequate or Imprecise Result or Readings	12/05/2023	12/05/2023
						F08	Hospitalization or Prolonged Hospitalization	A0908	Incorrect, Inadequate or Imprecise Result or Readings	12/05/2023	12/05/2023
						F1905	Device Revision or Replacement	A0908	Incorrect, Inadequate or Imprecise Result or Readings	12/05/2023	12/05/2023
					CT314	E2403	No Clinical Signs, Symptoms or Conditions	A0908	Incorrect, Inadequate or Imprecise Result or Readings	12/05/2023	12/05/2023
						F08	Hospitalization or Prolonged Hospitalization	A0908	Incorrect, Inadequate or Imprecise Result or Readings	12/05/2023	12/05/2023
						F1905	Device Revision or Replacement	A0908	Incorrect, Inadequate or Imprecise Result or Readings	12/05/2023	12/05/2023
					CT321	E2403	No Clinical Signs, Symptoms or Conditions	A0908	Incorrect, Inadequate or Imprecise Result or Readings	12/05/2023	12/05/2023
						F08	Hospitalization or Prolonged Hospitalization	A0908	Incorrect, Inadequate or Imprecise Result or Readings	12/05/2023	12/05/2023
						F1905	Device Revision or Replacement	A0908	Incorrect, Inadequate or Imprecise Result or Readings	12/05/2023	12/05/2023
					CT449	E2403	No Clinical Signs, Symptoms or Conditions	A0908	Incorrect, Inadequate or Imprecise Result or Readings	12/05/2023	12/05/2023

							F08	Hospitalization or Prolonged Hospitalization	A0908	Incorrect, Inadequate or Imprecise Result or Readings	12/05/2023	12/05/2023
							F1905	Device Revision or Replacement	A0908	Incorrect, Inadequate or Imprecise Result or Readings	12/05/2023	12/05/2023
					CT981	Single use	E2403	No Clinical Signs, Symptoms or Conditions	A0908	Incorrect, Inadequate or Imprecise Result or Readings	12/05/2023	12/05/2023
							F08	Hospitalization or Prolonged Hospitalization	A0908	Incorrect, Inadequate or Imprecise Result or Readings	12/05/2023	12/05/2023
							F1905	Device Revision or Replacement	A0908	Incorrect, Inadequate or Imprecise Result or Readings	12/05/2023	12/05/2023
					CT983	Surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A0908	Incorrect, Inadequate or Imprecise Result or Readings	12/05/2023	12/05/2023
							F08	Hospitalization or Prolonged Hospitalization	A0908	Incorrect, Inadequate or Imprecise Result or Readings	12/05/2023	12/05/2023
							F1905	Device Revision or Replacement	A0908	Incorrect, Inadequate or Imprecise Result or Readings	12/05/2023	12/05/2023
13807	2023	All other reportable incidents	37265	Dual-chamber implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A070909	Oversensing		
									A090801	Signal Artifact/Noise		
									A1301	Failure to Read Input Signal		
							F26	No Health Consequences or Impact	A070909	Oversensing		
									A090801	Signal Artifact/Noise		
									A1301	Failure to Read Input Signal		
13808	2023	All other reportable incidents	35852	Single-chamber implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A070909	Oversensing		
							F26	No Health Consequences or Impact	A070909	Oversensing		
13809	2023	All other reportable incidents	47265	Dual-chamber implantable pacemaker, rate-responsive	CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem	13/08/2024	13/08/2024
									A070504	Premature Discharge of Battery	13/08/2024	13/08/2024
							F08	Hospitalization or Prolonged Hospitalization	A0705	Battery Problem	13/08/2024	13/08/2024
									A070504	Premature Discharge of Battery	13/08/2024	13/08/2024
							F1905	Device Revision or Replacement	A0705	Battery Problem	13/08/2024	13/08/2024
									A070504	Premature Discharge of Battery	13/08/2024	13/08/2024
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem	13/08/2024	13/08/2024
									A070504	Premature Discharge of Battery	13/08/2024	13/08/2024
							F08	Hospitalization or Prolonged Hospitalization	A0705	Battery Problem	13/08/2024	13/08/2024
									A070504	Premature Discharge of Battery	13/08/2024	13/08/2024
							F1905	Device Revision or Replacement	A0705	Battery Problem	13/08/2024	13/08/2024
									A070504	Premature Discharge of Battery	13/08/2024	13/08/2024
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem	13/08/2024	13/08/2024
									A070504	Premature Discharge of Battery	13/08/2024	13/08/2024
							F08	Hospitalization or Prolonged Hospitalization	A0705	Battery Problem	13/08/2024	13/08/2024
									A070504	Premature Discharge of Battery	13/08/2024	13/08/2024
							F1905	Device Revision or Replacement	A0705	Battery Problem	13/08/2024	13/08/2024
									A070504	Premature Discharge of Battery	13/08/2024	13/08/2024
13810	2023	All other reportable incidents	47265	Dual-chamber implantable pacemaker, rate-responsive	CT141	Electrodes/Leads and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A0908	Incorrect, Inadequate or Imprecise Result or Readings	08/08/2023	08/08/2023
							F08	Hospitalization or Prolonged Hospitalization	A0908	Incorrect, Inadequate or Imprecise Result or Readings	08/08/2023	08/08/2023
							F1905	Device Revision or Replacement	A0908	Incorrect, Inadequate or Imprecise Result or Readings	08/08/2023	08/08/2023
					CT145	Cardiovascular devices	E2403	No Clinical Signs, Symptoms or Conditions	A0908	Incorrect, Inadequate or Imprecise Result or Readings	08/08/2023	08/08/2023
							F08	Hospitalization or Prolonged Hospitalization	A0908	Incorrect, Inadequate or Imprecise Result or Readings	08/08/2023	08/08/2023
							F1905	Device Revision or Replacement	A0908	Incorrect, Inadequate or Imprecise Result or Readings	08/08/2023	08/08/2023
					CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A0908	Incorrect, Inadequate or Imprecise Result or Readings	08/08/2023	08/08/2023
							F08	Hospitalization or Prolonged Hospitalization	A0908	Incorrect, Inadequate or Imprecise Result or Readings	08/08/2023	08/08/2023
							F1905	Device Revision or Replacement	A0908	Incorrect, Inadequate or Imprecise Result or Readings	08/08/2023	08/08/2023
					CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A0908	Incorrect, Inadequate or Imprecise Result or Readings	08/08/2023	08/08/2023
							F08	Hospitalization or Prolonged Hospitalization	A0908	Incorrect, Inadequate or Imprecise Result or Readings	08/08/2023	08/08/2023
							F1905	Device Revision or Replacement	A0908	Incorrect, Inadequate or Imprecise Result or Readings	08/08/2023	08/08/2023
					CT233	Surgical	E2403	No Clinical Signs, Symptoms or Conditions	A0908	Incorrect, Inadequate or Imprecise Result or Readings	08/08/2023	08/08/2023
							F08	Hospitalization or Prolonged Hospitalization	A0908	Incorrect, Inadequate or Imprecise Result or Readings	08/08/2023	08/08/2023
							F1905	Device Revision or Replacement	A0908	Incorrect, Inadequate or Imprecise Result or Readings	08/08/2023	08/08/2023
					CT2406	Implantable	E2403	No Clinical Signs, Symptoms or Conditions	A0908	Incorrect, Inadequate or Imprecise Result or Readings	08/08/2023	08/08/2023
							F08	Hospitalization or Prolonged Hospitalization	A0908	Incorrect, Inadequate or Imprecise Result or Readings	08/08/2023	08/08/2023
							F1905	Device Revision or Replacement	A0908	Incorrect, Inadequate or Imprecise Result or Readings	08/08/2023	08/08/2023
					CT2407	Active implantable	E2403	No Clinical Signs, Symptoms or Conditions	A0908	Incorrect, Inadequate or Imprecise Result or Readings	08/08/2023	08/08/2023
							F08	Hospitalization or Prolonged Hospitalization	A0908	Incorrect, Inadequate or Imprecise Result or Readings	08/08/2023	08/08/2023
							F1905	Device Revision or Replacement	A0908	Incorrect, Inadequate or Imprecise Result or Readings	08/08/2023	08/08/2023
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A0908	Incorrect, Inadequate or Imprecise Result or Readings	08/08/2023	08/08/2023
							F08	Hospitalization or Prolonged Hospitalization	A0908	Incorrect, Inadequate or Imprecise Result or Readings	08/08/2023	08/08/2023
							F1905	Device Revision or Replacement	A0908	Incorrect, Inadequate or Imprecise Result or Readings	08/08/2023	08/08/2023
					CT2727	Active implant devices/ancillaries	E2403	No Clinical Signs, Symptoms or Conditions	A0908	Incorrect, Inadequate or Imprecise Result or Readings	08/08/2023	08/08/2023
							F08	Hospitalization or Prolonged Hospitalization	A0908	Incorrect, Inadequate or Imprecise Result or Readings	08/08/2023	08/08/2023
							F1905	Device Revision or Replacement	A0908	Incorrect, Inadequate or Imprecise Result or Readings	08/08/2023	08/08/2023

					CT2750	Implantable cardiac electrical stimulators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A0908	Incorrect, Inadequate or Imprecise Result or Readings	08/08/2023	08/08/2023
							F08	Hospitalization or Prolonged Hospitalization	A0908	Incorrect, Inadequate or Imprecise Result or Readings	08/08/2023	08/08/2023
							F1905	Device Revision or Replacement	A0908	Incorrect, Inadequate or Imprecise Result or Readings	08/08/2023	08/08/2023
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A0908	Incorrect, Inadequate or Imprecise Result or Readings	08/08/2023	08/08/2023
							F08	Hospitalization or Prolonged Hospitalization	A0908	Incorrect, Inadequate or Imprecise Result or Readings	08/08/2023	08/08/2023
							F1905	Device Revision or Replacement	A0908	Incorrect, Inadequate or Imprecise Result or Readings	08/08/2023	08/08/2023
					CT2985	Active	E2403	No Clinical Signs, Symptoms or Conditions	A0908	Incorrect, Inadequate or Imprecise Result or Readings	08/08/2023	08/08/2023
							F08	Hospitalization or Prolonged Hospitalization	A0908	Incorrect, Inadequate or Imprecise Result or Readings	08/08/2023	08/08/2023
							F1905	Device Revision or Replacement	A0908	Incorrect, Inadequate or Imprecise Result or Readings	08/08/2023	08/08/2023
					CT3001	Implantable metal	E2403	No Clinical Signs, Symptoms or Conditions	A0908	Incorrect, Inadequate or Imprecise Result or Readings	08/08/2023	08/08/2023
							F08	Hospitalization or Prolonged Hospitalization	A0908	Incorrect, Inadequate or Imprecise Result or Readings	08/08/2023	08/08/2023
							F1905	Device Revision or Replacement	A0908	Incorrect, Inadequate or Imprecise Result or Readings	08/08/2023	08/08/2023
					CT3002	Implantable synthetic polymer	E2403	No Clinical Signs, Symptoms or Conditions	A0908	Incorrect, Inadequate or Imprecise Result or Readings	08/08/2023	08/08/2023
							F08	Hospitalization or Prolonged Hospitalization	A0908	Incorrect, Inadequate or Imprecise Result or Readings	08/08/2023	08/08/2023
							F1905	Device Revision or Replacement	A0908	Incorrect, Inadequate or Imprecise Result or Readings	08/08/2023	08/08/2023
					CT3006	Implantable inorganic material	E2403	No Clinical Signs, Symptoms or Conditions	A0908	Incorrect, Inadequate or Imprecise Result or Readings	08/08/2023	08/08/2023
							F08	Hospitalization or Prolonged Hospitalization	A0908	Incorrect, Inadequate or Imprecise Result or Readings	08/08/2023	08/08/2023
							F1905	Device Revision or Replacement	A0908	Incorrect, Inadequate or Imprecise Result or Readings	08/08/2023	08/08/2023
					CT314	Electrically powered	E2403	No Clinical Signs, Symptoms or Conditions	A0908	Incorrect, Inadequate or Imprecise Result or Readings	08/08/2023	08/08/2023
							F08	Hospitalization or Prolonged Hospitalization	A0908	Incorrect, Inadequate or Imprecise Result or Readings	08/08/2023	08/08/2023
							F1905	Device Revision or Replacement	A0908	Incorrect, Inadequate or Imprecise Result or Readings	08/08/2023	08/08/2023
					CT321	Long-term surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A0908	Incorrect, Inadequate or Imprecise Result or Readings	08/08/2023	08/08/2023
							F08	Hospitalization or Prolonged Hospitalization	A0908	Incorrect, Inadequate or Imprecise Result or Readings	08/08/2023	08/08/2023
							F1905	Device Revision or Replacement	A0908	Incorrect, Inadequate or Imprecise Result or Readings	08/08/2023	08/08/2023
					CT449	Generators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A0908	Incorrect, Inadequate or Imprecise Result or Readings	08/08/2023	08/08/2023
							F08	Hospitalization or Prolonged Hospitalization	A0908	Incorrect, Inadequate or Imprecise Result or Readings	08/08/2023	08/08/2023
							F1905	Device Revision or Replacement	A0908	Incorrect, Inadequate or Imprecise Result or Readings	08/08/2023	08/08/2023
					CT981	Single use	E2403	No Clinical Signs, Symptoms or Conditions	A0908	Incorrect, Inadequate or Imprecise Result or Readings	08/08/2023	08/08/2023
							F08	Hospitalization or Prolonged Hospitalization	A0908	Incorrect, Inadequate or Imprecise Result or Readings	08/08/2023	08/08/2023
							F1905	Device Revision or Replacement	A0908	Incorrect, Inadequate or Imprecise Result or Readings	08/08/2023	08/08/2023
					CT983	Surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A0908	Incorrect, Inadequate or Imprecise Result or Readings	08/08/2023	08/08/2023
							F08	Hospitalization or Prolonged Hospitalization	A0908	Incorrect, Inadequate or Imprecise Result or Readings	08/08/2023	08/08/2023
							F1905	Device Revision or Replacement	A0908	Incorrect, Inadequate or Imprecise Result or Readings	08/08/2023	08/08/2023
13811	2023	All other reportable incidents	47263	Cardiac resynchronization therapy implantable pacemaker	CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A0908	Incorrect, Inadequate or Imprecise Result or Readings	17/03/2023	17/03/2023
							F08	Hospitalization or Prolonged Hospitalization	A0908	Incorrect, Inadequate or Imprecise Result or Readings	17/03/2023	17/03/2023
							F1905	Device Revision or Replacement	A0908	Incorrect, Inadequate or Imprecise Result or Readings	17/03/2023	17/03/2023
					CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A0908	Incorrect, Inadequate or Imprecise Result or Readings	17/03/2023	17/03/2023
							F08	Hospitalization or Prolonged Hospitalization	A0908	Incorrect, Inadequate or Imprecise Result or Readings	17/03/2023	17/03/2023
							F1905	Device Revision or Replacement	A0908	Incorrect, Inadequate or Imprecise Result or Readings	17/03/2023	17/03/2023
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A0908	Incorrect, Inadequate or Imprecise Result or Readings	17/03/2023	17/03/2023
							F08	Hospitalization or Prolonged Hospitalization	A0908	Incorrect, Inadequate or Imprecise Result or Readings	17/03/2023	17/03/2023
							F1905	Device Revision or Replacement	A0908	Incorrect, Inadequate or Imprecise Result or Readings	17/03/2023	17/03/2023
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A0908	Incorrect, Inadequate or Imprecise Result or Readings	17/03/2023	17/03/2023
							F08	Hospitalization or Prolonged Hospitalization	A0908	Incorrect, Inadequate or Imprecise Result or Readings	17/03/2023	17/03/2023
							F1905	Device Revision or Replacement	A0908	Incorrect, Inadequate or Imprecise Result or Readings	17/03/2023	17/03/2023
13812	2023	All other reportable incidents	17882	Physiologic-monitoring defibrillation system	CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A1603	Failure of Device to Self-Test		
							F27	Problem Identified During Non-Clinical Procedure	A1603	Failure of Device to Self-Test		
13813	2023	All other reportable incidents	47270	Cardiac resynchronization therapy implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	04/05/2023	04/05/2023
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	04/05/2023	04/05/2023
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	04/05/2023	04/05/2023
13814	2023	All other reportable incidents	47205	Cardiac pulse generator programmer	CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A150105	Key or Button Unresponsive/not Working		
							F26	No Health Consequences or Impact	A150105	Key or Button Unresponsive/not Working		
13815	2023	All other reportable incidents	35852	Single-chamber implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E2104	Electric Shock	A070909	Oversensing		

									A071302	Inappropriate/Inadequate Shock/Stimulation		
									A090801	Signal Artifact/Noise		
							F08	Hospitalization or Prolonged Hospitalization	A070909	Oversensing		
									A071302	Inappropriate/Inadequate Shock/Stimulation		
									A090801	Signal Artifact/Noise		
							F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A070909	Oversensing		
									A071302	Inappropriate/Inadequate Shock/Stimulation		
									A090801	Signal Artifact/Noise		
13816	2023	All other reportable incidents	47263	Cardiac resynchronization therapy implantable pacemaker	CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A0908	Incorrect, Inadequate or Imprecise Result or Readings	08/08/2023	08/08/2023
							F08	Hospitalization or Prolonged Hospitalization	A0908	Incorrect, Inadequate or Imprecise Result or Readings	08/08/2023	08/08/2023
							F1905	Device Revision or Replacement	A0908	Incorrect, Inadequate or Imprecise Result or Readings	08/08/2023	08/08/2023
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A0908	Incorrect, Inadequate or Imprecise Result or Readings	08/08/2023	08/08/2023
							F08	Hospitalization or Prolonged Hospitalization	A0908	Incorrect, Inadequate or Imprecise Result or Readings	08/08/2023	08/08/2023
							F1905	Device Revision or Replacement	A0908	Incorrect, Inadequate or Imprecise Result or Readings	08/08/2023	08/08/2023
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A0908	Incorrect, Inadequate or Imprecise Result or Readings	08/08/2023	08/08/2023
							F08	Hospitalization or Prolonged Hospitalization	A0908	Incorrect, Inadequate or Imprecise Result or Readings	08/08/2023	08/08/2023
							F1905	Device Revision or Replacement	A0908	Incorrect, Inadequate or Imprecise Result or Readings	08/08/2023	08/08/2023
13817	2023	All other reportable incidents	36241	Pacemaker/defibrillator lead	CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A072201	High Impedance		
							F26	No Health Consequences or Impact	A072201	High Impedance		
13818	2023	All other reportable incidents	47265	Dual-chamber implantable pacemaker, rate-responsive	CT1685	Cardiac pacemakers	E011903	Syncope/Fainting	A070909	Oversensing	03/08/2023	03/08/2023
									A0712	Pacing Problem	03/08/2023	03/08/2023
									A0908	Incorrect, Inadequate or Imprecise Result or Readings	03/08/2023	03/08/2023
							F08	Hospitalization or Prolonged Hospitalization	A070909	Oversensing	03/08/2023	03/08/2023
									A0712	Pacing Problem	03/08/2023	03/08/2023
									A0908	Incorrect, Inadequate or Imprecise Result or Readings	03/08/2023	03/08/2023
							F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A070909	Oversensing	03/08/2023	03/08/2023
									A0712	Pacing Problem	03/08/2023	03/08/2023
									A0908	Incorrect, Inadequate or Imprecise Result or Readings	03/08/2023	03/08/2023
							F1001	Absence of Treatment	A070909	Oversensing	03/08/2023	03/08/2023
									A0712	Pacing Problem	03/08/2023	03/08/2023
									A0908	Incorrect, Inadequate or Imprecise Result or Readings	03/08/2023	03/08/2023
							F15	Recognized Device or Procedural Complication	A070909	Oversensing	03/08/2023	03/08/2023
									A0712	Pacing Problem	03/08/2023	03/08/2023
									A0908	Incorrect, Inadequate or Imprecise Result or Readings	03/08/2023	03/08/2023
							F1905	Device Revision or Replacement	A070909	Oversensing	03/08/2023	03/08/2023
									A0712	Pacing Problem	03/08/2023	03/08/2023
									A0908	Incorrect, Inadequate or Imprecise Result or Readings	03/08/2023	03/08/2023
					CT257	Implantable cardiac pacemakers	E011903	Syncope/Fainting	A070909	Oversensing	03/08/2023	03/08/2023
									A0712	Pacing Problem	03/08/2023	03/08/2023
									A0908	Incorrect, Inadequate or Imprecise Result or Readings	03/08/2023	03/08/2023
							F08	Hospitalization or Prolonged Hospitalization	A070909	Oversensing	03/08/2023	03/08/2023
									A0712	Pacing Problem	03/08/2023	03/08/2023
									A0908	Incorrect, Inadequate or Imprecise Result or Readings	03/08/2023	03/08/2023
							F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A070909	Oversensing	03/08/2023	03/08/2023
									A0712	Pacing Problem	03/08/2023	03/08/2023
									A0908	Incorrect, Inadequate or Imprecise Result or Readings	03/08/2023	03/08/2023
							F1001	Absence of Treatment	A070909	Oversensing	03/08/2023	03/08/2023
									A0712	Pacing Problem	03/08/2023	03/08/2023
									A0908	Incorrect, Inadequate or Imprecise Result or Readings	03/08/2023	03/08/2023
							F15	Recognized Device or Procedural Complication	A070909	Oversensing	03/08/2023	03/08/2023
									A0712	Pacing Problem	03/08/2023	03/08/2023
									A0908	Incorrect, Inadequate or Imprecise Result or Readings	03/08/2023	03/08/2023
							F1905	Device Revision or Replacement	A070909	Oversensing	03/08/2023	03/08/2023
									A0712	Pacing Problem	03/08/2023	03/08/2023
									A0908	Incorrect, Inadequate or Imprecise Result or Readings	03/08/2023	03/08/2023
					CT2866	Implantable cardiac electrical stimulators	E011903	Syncope/Fainting	A070909	Oversensing	03/08/2023	03/08/2023
									A0712	Pacing Problem	03/08/2023	03/08/2023
									A0908	Incorrect, Inadequate or Imprecise Result or Readings	03/08/2023	03/08/2023
							F08	Hospitalization or Prolonged Hospitalization	A070909	Oversensing	03/08/2023	03/08/2023
									A0712	Pacing Problem	03/08/2023	03/08/2023
									A0908	Incorrect, Inadequate or Imprecise Result or Readings	03/08/2023	03/08/2023
							F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A070909	Oversensing	03/08/2023	03/08/2023
									A0712	Pacing Problem	03/08/2023	03/08/2023
									A0908	Incorrect, Inadequate or Imprecise Result or Readings	03/08/2023	03/08/2023
							F1001	Absence of Treatment	A070909	Oversensing	03/08/2023	03/08/2023
									A0712	Pacing Problem	03/08/2023	03/08/2023
									A0908	Incorrect, Inadequate or Imprecise Result or Readings	03/08/2023	03/08/2023
							F15	Recognized Device or Procedural Complication	A070909	Oversensing	03/08/2023	03/08/2023
									A0712	Pacing Problem	03/08/2023	03/08/2023

								A0908	Incorrect, Inadequate or Imprecise Result or Readings	03/08/2023	03/08/2023
						F1905	Device Revision or Replacement	A070909	Oversensing	03/08/2023	03/08/2023
								A0712	Pacing Problem	03/08/2023	03/08/2023
								A0908	Incorrect, Inadequate or Imprecise Result or Readings	03/08/2023	03/08/2023
13819	2023	All other reportable incidents	47265	Dual-chamber implantable pacemaker, rate-responsive	CT141	Electrodes/Leads and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT145	Cardiovascular devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT233	Surgical	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT2406	Implantable	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT2407	Active implantable	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT2727	Active implant devices/ancillaries	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT2750	Implantable cardiac electrical stimulators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT2985	Active	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT3001	Implantable metal	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT3002	Implantable synthetic polymer	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT3006	Implantable inorganic material	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT314	Electrically powered	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT321	Long-term surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT449	Generators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT981	Single use	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT983	Surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
13820	2023	All other reportable incidents	47270	Cardiac resynchronization therapy implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
13821	2023	All other reportable incidents	47265	Dual-chamber implantable pacemaker, rate-responsive	CT141	Electrodes/Leads and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	12/05/2023	12/05/2023
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	12/05/2023	12/05/2023
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	12/05/2023	12/05/2023
					CT145	Cardiovascular devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	12/05/2023	12/05/2023
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	12/05/2023	12/05/2023
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	12/05/2023	12/05/2023
					CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	12/05/2023	12/05/2023
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	12/05/2023	12/05/2023
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	12/05/2023	12/05/2023
					CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	12/05/2023	12/05/2023
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	12/05/2023	12/05/2023
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	12/05/2023	12/05/2023
					CT233	Surgical	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	12/05/2023	12/05/2023
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	12/05/2023	12/05/2023
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	12/05/2023	12/05/2023
					CT2406	Implantable	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	12/05/2023	12/05/2023
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	12/05/2023	12/05/2023
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	12/05/2023	12/05/2023
					CT2407	Active implantable	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	12/05/2023	12/05/2023
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	12/05/2023	12/05/2023
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	12/05/2023	12/05/2023
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	12/05/2023	12/05/2023
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	12/05/2023	12/05/2023
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	12/05/2023	12/05/2023
					CT2727	Active implant devices/ancillaries	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	12/05/2023	12/05/2023
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	12/05/2023	12/05/2023
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	12/05/2023	12/05/2023
					CT2750	Implantable cardiac electrical stimulators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	12/05/2023	12/05/2023
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	12/05/2023	12/05/2023
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	12/05/2023	12/05/2023
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	12/05/2023	12/05/2023
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	12/05/2023	12/05/2023
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	12/05/2023	12/05/2023
					CT2985	Active	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	12/05/2023	12/05/2023
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	12/05/2023	12/05/2023
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	12/05/2023	12/05/2023
					CT3001	Implantable metal	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	12/05/2023	12/05/2023
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	12/05/2023	12/05/2023
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	12/05/2023	12/05/2023
					CT3002	Implantable synthetic polymer	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	12/05/2023	12/05/2023

						F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	12/05/2023	12/05/2023
						F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	12/05/2023	12/05/2023
					CT3006	Implantable inorganic material	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	12/05/2023	12/05/2023
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	12/05/2023	12/05/2023
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	12/05/2023	12/05/2023
					CT314	Electrically powered	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	12/05/2023	12/05/2023
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	12/05/2023	12/05/2023
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	12/05/2023	12/05/2023
					CT321	Long-term surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	12/05/2023	12/05/2023
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	12/05/2023	12/05/2023
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	12/05/2023	12/05/2023
					CT449	Generators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	12/05/2023	12/05/2023
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	12/05/2023	12/05/2023
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	12/05/2023	12/05/2023
					CT981	Single use	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	12/05/2023	12/05/2023
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	12/05/2023	12/05/2023
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	12/05/2023	12/05/2023
					CT983	Surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	12/05/2023	12/05/2023
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	12/05/2023	12/05/2023
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	12/05/2023	12/05/2023
13822	2023	All other reportable incidents	47265	Dual-chamber implantable pacemaker, rate-responsive	CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A070909	Oversensing		
									A1107	Data Problem		
							F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A070909	Oversensing		
									A1107	Data Problem		
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A070909	Oversensing		
									A1107	Data Problem		
							F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A070909	Oversensing		
									A1107	Data Problem		
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A070909	Oversensing		
									A1107	Data Problem		
							F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A070909	Oversensing		
									A1107	Data Problem		
13823	2023		47265	Dual-chamber implantable pacemaker, rate-responsive	CT141	Electrodes/Leads and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	17/06/2023	17/06/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	17/06/2023	17/06/2023
					CT145	Cardiovascular devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	17/06/2023	17/06/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	17/06/2023	17/06/2023
					CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	17/06/2023	17/06/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	17/06/2023	17/06/2023
					CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	17/06/2023	17/06/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	17/06/2023	17/06/2023
					CT233	Surgical	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	17/06/2023	17/06/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	17/06/2023	17/06/2023
					CT2406	Implantable	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	17/06/2023	17/06/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	17/06/2023	17/06/2023
					CT2407	Active implantable	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	17/06/2023	17/06/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	17/06/2023	17/06/2023
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	17/06/2023	17/06/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	17/06/2023	17/06/2023
					CT2727	Active implant devices/ancillaries	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	17/06/2023	17/06/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	17/06/2023	17/06/2023
					CT2750	Implantable cardiac electrical stimulators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	17/06/2023	17/06/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	17/06/2023	17/06/2023
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	17/06/2023	17/06/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	17/06/2023	17/06/2023
					CT2985	Active	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	17/06/2023	17/06/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	17/06/2023	17/06/2023
					CT3001	Implantable metal	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	17/06/2023	17/06/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	17/06/2023	17/06/2023
					CT3002	Implantable synthetic polymer	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	17/06/2023	17/06/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	17/06/2023	17/06/2023
					CT3006	Implantable inorganic material	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	17/06/2023	17/06/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	17/06/2023	17/06/2023
					CT314	Electrically powered	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	17/06/2023	17/06/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	17/06/2023	17/06/2023
					CT321	Long-term surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	17/06/2023	17/06/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	17/06/2023	17/06/2023
					CT449	Generators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	17/06/2023	17/06/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	17/06/2023	17/06/2023
					CT981	Single use	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	17/06/2023	17/06/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	17/06/2023	17/06/2023
					CT983	Surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	17/06/2023	17/06/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	17/06/2023	17/06/2023
13824	2023	All other reportable incidents	47205	Cardiac pulse generator programmer	CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A0203	Defective Device		
									A150105	Key or Button Unresponsive/not Working		
							F26	No Health Consequences or Impact	A0203	Defective Device		
									A150105	Key or Button Unresponsive/not Working		
13825	2023	All other reportable incidents	47265	Dual-chamber implantable pacemaker, rate-responsive	CT141	Electrodes/Leads and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070909	Oversensing		
									A070910	Undersensing		
									A0712	Pacing Problem		
							F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A070909	Oversensing		
									A070910	Undersensing		
									A0712	Pacing Problem		
					CT145	Cardiovascular devices	E2403	No Clinical Signs, Symptoms or Conditions	A070909	Oversensing		
									A070910	Undersensing		
									A0712	Pacing Problem		
							F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A070909	Oversensing		
									A070910	Undersensing		
									A0712	Pacing Problem		

					CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070909	Oversensing		
									A070910	Undersensing		
									A0712	Pacing Problem		
							F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A070909	Oversensing		
									A070910	Undersensing		
									A0712	Pacing Problem		
					CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A070909	Oversensing		
									A070910	Undersensing		
									A0712	Pacing Problem		
							F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A070909	Oversensing		
									A070910	Undersensing		
									A0712	Pacing Problem		
					CT233	Surgical	E2403	No Clinical Signs, Symptoms or Conditions	A070909	Oversensing		
									A070910	Undersensing		
									A0712	Pacing Problem		
							F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A070909	Oversensing		
									A070910	Undersensing		
									A0712	Pacing Problem		
					CT2406	Implantable	E2403	No Clinical Signs, Symptoms or Conditions	A070909	Oversensing		
									A070910	Undersensing		
									A0712	Pacing Problem		
							F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A070909	Oversensing		
									A070910	Undersensing		
									A0712	Pacing Problem		
					CT2407	Active implantable	E2403	No Clinical Signs, Symptoms or Conditions	A070909	Oversensing		
									A070910	Undersensing		
									A0712	Pacing Problem		
							F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A070909	Oversensing		
									A070910	Undersensing		
									A0712	Pacing Problem		
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A070909	Oversensing		
									A070910	Undersensing		
									A0712	Pacing Problem		
							F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A070909	Oversensing		
									A070910	Undersensing		
									A0712	Pacing Problem		
					CT2727	Active implant devices/ancillaries	E2403	No Clinical Signs, Symptoms or Conditions	A070909	Oversensing		
									A070910	Undersensing		
									A0712	Pacing Problem		
							F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A070909	Oversensing		
									A070910	Undersensing		
									A0712	Pacing Problem		
					CT2750	Implantable cardiac electrical stimulators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070909	Oversensing		
									A070910	Undersensing		
									A0712	Pacing Problem		
							F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A070909	Oversensing		
									A070910	Undersensing		
									A0712	Pacing Problem		
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A070909	Oversensing		
									A070910	Undersensing		
									A0712	Pacing Problem		
							F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A070909	Oversensing		
									A070910	Undersensing		
									A0712	Pacing Problem		
					CT2985	Active	E2403	No Clinical Signs, Symptoms or Conditions	A070909	Oversensing		
									A070910	Undersensing		
									A0712	Pacing Problem		
							F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A070909	Oversensing		
									A070910	Undersensing		
									A0712	Pacing Problem		
					CT3001	Implantable metal	E2403	No Clinical Signs, Symptoms or Conditions	A070909	Oversensing		
									A070910	Undersensing		
									A0712	Pacing Problem		
							F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A070909	Oversensing		
									A070910	Undersensing		
									A0712	Pacing Problem		
					CT3002	Implantable synthetic polymer	E2403	No Clinical Signs, Symptoms or Conditions	A070909	Oversensing		
									A070910	Undersensing		
									A0712	Pacing Problem		
							F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A070909	Oversensing		
									A070910	Undersensing		
									A0712	Pacing Problem		
					CT3006	Implantable inorganic material	E2403	No Clinical Signs, Symptoms or Conditions	A070909	Oversensing		
									A070910	Undersensing		
									A0712	Pacing Problem		
							F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A070909	Oversensing		
									A070910	Undersensing		
									A0712	Pacing Problem		
					CT314	Electrically powered	E2403	No Clinical Signs, Symptoms or Conditions	A070909	Oversensing		
									A070910	Undersensing		
									A0712	Pacing Problem		
							F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A070909	Oversensing		
									A070910	Undersensing		
									A0712	Pacing Problem		
					CT321	Long-term surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A070909	Oversensing		
									A070910	Undersensing		
									A0712	Pacing Problem		
							F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A070909	Oversensing		

									A070910	Undersensing		
									A0712	Pacing Problem		
					CT449	Generators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070909	Oversensing		
									A070910	Undersensing		
									A0712	Pacing Problem		
							F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A070909	Oversensing		
									A070910	Undersensing		
									A0712	Pacing Problem		
					CT981	Single use	E2403	No Clinical Signs, Symptoms or Conditions	A070909	Oversensing		
									A070910	Undersensing		
									A0712	Pacing Problem		
							F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A070909	Oversensing		
									A070910	Undersensing		
									A0712	Pacing Problem		
					CT983	Surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A070909	Oversensing		
									A070910	Undersensing		
									A0712	Pacing Problem		
							F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A070909	Oversensing		
									A070910	Undersensing		
									A0712	Pacing Problem		
13826	2023	All other reportable incidents	37265	Dual-chamber implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	04/04/2023	04/04/2023
							F1901	Additional Surgery	A070504	Premature Discharge of Battery	04/04/2023	04/04/2023
13827	2023	All other reportable incidents	35223	Endocardial/interventricular septal pacing lead	CT1684	Cardiac pacemakers and associated devices	E060106	Heart Block	A070909	Oversensing		
							F12	Serious Injury/ Illness/ Impairment	A070909	Oversensing		
							F23	Unexpected Medical Intervention	A070909	Oversensing		
13828	2023	All other reportable incidents	35223	Endocardial/interventricular septal pacing lead	CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A07	Electrical /Electronic Property Problem		
									A072201	High Impedance		
									A090801	Signal Artifact/Noise		
							F12	Serious Injury/ Illness/ Impairment	A07	Electrical /Electronic Property Problem		
									A072201	High Impedance		
									A090801	Signal Artifact/Noise		
							F23	Unexpected Medical Intervention	A07	Electrical /Electronic Property Problem		
									A072201	High Impedance		
									A090801	Signal Artifact/Noise		
13829	2023	All other reportable incidents	35223	Endocardial/interventricular septal pacing lead	CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070104	Unstable Capture Threshold		
									A070909	Oversensing		
									A0722	Impedance Problem		
									A090801	Signal Artifact/Noise		
							F26	No Health Consequences or Impact	A070104	Unstable Capture Threshold		
									A070909	Oversensing		
									A0722	Impedance Problem		
									A090801	Signal Artifact/Noise		
13830	2023	All other reportable incidents	35223	Endocardial/interventricular septal pacing lead	CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A072201	High Impedance		
							F26	No Health Consequences or Impact	A072201	High Impedance		
13831	2023	All other reportable incidents	60789	Intracardiac pacemaker	CT130	Catheters and associated devices	E060106	Heart Block	A0701	Capturing Problem	24/05/2023	24/05/2023
									A070101	Failure to Capture	24/05/2023	24/05/2023
									A070102	High Capture Threshold	24/05/2023	24/05/2023
									A070104	Unstable Capture Threshold	24/05/2023	24/05/2023
							F08	Hospitalization or Prolonged Hospitalization	A0701	Capturing Problem	24/05/2023	24/05/2023
									A070101	Failure to Capture	24/05/2023	24/05/2023
									A070102	High Capture Threshold	24/05/2023	24/05/2023
									A070104	Unstable Capture Threshold	24/05/2023	24/05/2023
							F12	Serious Injury/ Illness/ Impairment	A0701	Capturing Problem	24/05/2023	24/05/2023
									A070101	Failure to Capture	24/05/2023	24/05/2023
									A070102	High Capture Threshold	24/05/2023	24/05/2023
									A070104	Unstable Capture Threshold	24/05/2023	24/05/2023
							F1905	Device Revision or Replacement	A0701	Capturing Problem	24/05/2023	24/05/2023
									A070101	Failure to Capture	24/05/2023	24/05/2023
									A070102	High Capture Threshold	24/05/2023	24/05/2023
									A070104	Unstable Capture Threshold	24/05/2023	24/05/2023
					CT145	Cardiovascular devices	E060106	Heart Block	A0701	Capturing Problem	24/05/2023	24/05/2023
									A070101	Failure to Capture	24/05/2023	24/05/2023
									A070102	High Capture Threshold	24/05/2023	24/05/2023
									A070104	Unstable Capture Threshold	24/05/2023	24/05/2023
							F08	Hospitalization or Prolonged Hospitalization	A0701	Capturing Problem	24/05/2023	24/05/2023
									A070101	Failure to Capture	24/05/2023	24/05/2023
									A070102	High Capture Threshold	24/05/2023	24/05/2023
									A070104	Unstable Capture Threshold	24/05/2023	24/05/2023
							F12	Serious Injury/ Illness/ Impairment	A0701	Capturing Problem	24/05/2023	24/05/2023
									A070101	Failure to Capture	24/05/2023	24/05/2023
									A070102	High Capture Threshold	24/05/2023	24/05/2023
									A070104	Unstable Capture Threshold	24/05/2023	24/05/2023
							F1905	Device Revision or Replacement	A0701	Capturing Problem	24/05/2023	24/05/2023
									A070101	Failure to Capture	24/05/2023	24/05/2023
									A070102	High Capture Threshold	24/05/2023	24/05/2023
									A070104	Unstable Capture Threshold	24/05/2023	24/05/2023
					CT1684	Cardiac pacemakers and associated devices	E060106	Heart Block	A0701	Capturing Problem	24/05/2023	24/05/2023
									A070101	Failure to Capture	24/05/2023	24/05/2023
									A070102	High Capture Threshold	24/05/2023	24/05/2023
									A070104	Unstable Capture Threshold	24/05/2023	24/05/2023
							F08	Hospitalization or Prolonged Hospitalization	A0701	Capturing Problem	24/05/2023	24/05/2023
									A070101	Failure to Capture	24/05/2023	24/05/2023
									A070102	High Capture Threshold	24/05/2023	24/05/2023
									A070104	Unstable Capture Threshold	24/05/2023	24/05/2023
							F12	Serious Injury/ Illness/ Impairment	A0701	Capturing Problem	24/05/2023	24/05/2023
									A070101	Failure to Capture	24/05/2023	24/05/2023
									A070102	High Capture Threshold	24/05/2023	24/05/2023
									A070104	Unstable Capture Threshold	24/05/2023	24/05/2023

							F1905	Device Revision or Replacement	A0701	Capturing Problem	24/05/2023	24/05/2023
									A070101	Failure to Capture	24/05/2023	24/05/2023
									A070102	High Capture Threshold	24/05/2023	24/05/2023
									A070104	Unstable Capture Threshold	24/05/2023	24/05/2023
					CT1685	Cardiac pacemakers	E060106	Heart Block	A0701	Capturing Problem	24/05/2023	24/05/2023
									A070101	Failure to Capture	24/05/2023	24/05/2023
									A070102	High Capture Threshold	24/05/2023	24/05/2023
									A070104	Unstable Capture Threshold	24/05/2023	24/05/2023
							F08	Hospitalization or Prolonged Hospitalization	A0701	Capturing Problem	24/05/2023	24/05/2023
									A070101	Failure to Capture	24/05/2023	24/05/2023
									A070102	High Capture Threshold	24/05/2023	24/05/2023
									A070104	Unstable Capture Threshold	24/05/2023	24/05/2023
							F12	Serious Injury/ Illness/ Impairment	A0701	Capturing Problem	24/05/2023	24/05/2023
									A070101	Failure to Capture	24/05/2023	24/05/2023
									A070102	High Capture Threshold	24/05/2023	24/05/2023
									A070104	Unstable Capture Threshold	24/05/2023	24/05/2023
							F1905	Device Revision or Replacement	A0701	Capturing Problem	24/05/2023	24/05/2023
									A070101	Failure to Capture	24/05/2023	24/05/2023
									A070102	High Capture Threshold	24/05/2023	24/05/2023
									A070104	Unstable Capture Threshold	24/05/2023	24/05/2023
					CT233	Surgical	E060106	Heart Block	A0701	Capturing Problem	24/05/2023	24/05/2023
									A070101	Failure to Capture	24/05/2023	24/05/2023
									A070102	High Capture Threshold	24/05/2023	24/05/2023
									A070104	Unstable Capture Threshold	24/05/2023	24/05/2023
							F08	Hospitalization or Prolonged Hospitalization	A0701	Capturing Problem	24/05/2023	24/05/2023
									A070101	Failure to Capture	24/05/2023	24/05/2023
									A070102	High Capture Threshold	24/05/2023	24/05/2023
									A070104	Unstable Capture Threshold	24/05/2023	24/05/2023
							F12	Serious Injury/ Illness/ Impairment	A0701	Capturing Problem	24/05/2023	24/05/2023
									A070101	Failure to Capture	24/05/2023	24/05/2023
									A070102	High Capture Threshold	24/05/2023	24/05/2023
									A070104	Unstable Capture Threshold	24/05/2023	24/05/2023
							F1905	Device Revision or Replacement	A0701	Capturing Problem	24/05/2023	24/05/2023
									A070101	Failure to Capture	24/05/2023	24/05/2023
									A070102	High Capture Threshold	24/05/2023	24/05/2023
									A070104	Unstable Capture Threshold	24/05/2023	24/05/2023
					CT2406	Implantable	E060106	Heart Block	A0701	Capturing Problem	24/05/2023	24/05/2023
									A070101	Failure to Capture	24/05/2023	24/05/2023
									A070102	High Capture Threshold	24/05/2023	24/05/2023
									A070104	Unstable Capture Threshold	24/05/2023	24/05/2023
							F08	Hospitalization or Prolonged Hospitalization	A0701	Capturing Problem	24/05/2023	24/05/2023
									A070101	Failure to Capture	24/05/2023	24/05/2023
									A070102	High Capture Threshold	24/05/2023	24/05/2023
									A070104	Unstable Capture Threshold	24/05/2023	24/05/2023
							F12	Serious Injury/ Illness/ Impairment	A0701	Capturing Problem	24/05/2023	24/05/2023
									A070101	Failure to Capture	24/05/2023	24/05/2023
									A070102	High Capture Threshold	24/05/2023	24/05/2023
									A070104	Unstable Capture Threshold	24/05/2023	24/05/2023
							F1905	Device Revision or Replacement	A0701	Capturing Problem	24/05/2023	24/05/2023
									A070101	Failure to Capture	24/05/2023	24/05/2023
									A070102	High Capture Threshold	24/05/2023	24/05/2023
									A070104	Unstable Capture Threshold	24/05/2023	24/05/2023
					CT2407	Active implantable	E060106	Heart Block	A0701	Capturing Problem	24/05/2023	24/05/2023
									A070101	Failure to Capture	24/05/2023	24/05/2023
									A070102	High Capture Threshold	24/05/2023	24/05/2023
									A070104	Unstable Capture Threshold	24/05/2023	24/05/2023
							F08	Hospitalization or Prolonged Hospitalization	A0701	Capturing Problem	24/05/2023	24/05/2023
									A070101	Failure to Capture	24/05/2023	24/05/2023
									A070102	High Capture Threshold	24/05/2023	24/05/2023
									A070104	Unstable Capture Threshold	24/05/2023	24/05/2023
							F12	Serious Injury/ Illness/ Impairment	A0701	Capturing Problem	24/05/2023	24/05/2023
									A070101	Failure to Capture	24/05/2023	24/05/2023
									A070102	High Capture Threshold	24/05/2023	24/05/2023
									A070104	Unstable Capture Threshold	24/05/2023	24/05/2023
							F1905	Device Revision or Replacement	A0701	Capturing Problem	24/05/2023	24/05/2023
									A070101	Failure to Capture	24/05/2023	24/05/2023
									A070102	High Capture Threshold	24/05/2023	24/05/2023
									A070104	Unstable Capture Threshold	24/05/2023	24/05/2023
					CT257	Implantable cardiac pacemakers	E060106	Heart Block	A0701	Capturing Problem	24/05/2023	24/05/2023
									A070101	Failure to Capture	24/05/2023	24/05/2023
									A070102	High Capture Threshold	24/05/2023	24/05/2023
									A070104	Unstable Capture Threshold	24/05/2023	24/05/2023
							F08	Hospitalization or Prolonged Hospitalization	A0701	Capturing Problem	24/05/2023	24/05/2023
									A070101	Failure to Capture	24/05/2023	24/05/2023
									A070102	High Capture Threshold	24/05/2023	24/05/2023
									A070104	Unstable Capture Threshold	24/05/2023	24/05/2023
							F12	Serious Injury/ Illness/ Impairment	A0701	Capturing Problem	24/05/2023	24/05/2023
									A070101	Failure to Capture	24/05/2023	24/05/2023
									A070102	High Capture Threshold	24/05/2023	24/05/2023
									A070104	Unstable Capture Threshold	24/05/2023	24/05/2023
							F1905	Device Revision or Replacement	A0701	Capturing Problem	24/05/2023	24/05/2023
									A070101	Failure to Capture	24/05/2023	24/05/2023
									A070102	High Capture Threshold	24/05/2023	24/05/2023
									A070104	Unstable Capture Threshold	24/05/2023	24/05/2023
					CT2727	Active implant devices/ancillaries	E060106	Heart Block	A0701	Capturing Problem	24/05/2023	24/05/2023
									A070101	Failure to Capture	24/05/2023	24/05/2023
									A070102	High Capture Threshold	24/05/2023	24/05/2023
									A070104	Unstable Capture Threshold	24/05/2023	24/05/2023

							F08	Hospitalization or Prolonged Hospitalization	A0701	Capturing Problem	24/05/2023	24/05/2023
									A070101	Failure to Capture	24/05/2023	24/05/2023
									A070102	High Capture Threshold	24/05/2023	24/05/2023
									A070104	Unstable Capture Threshold	24/05/2023	24/05/2023
							F12	Serious Injury/ Illness/ Impairment	A0701	Capturing Problem	24/05/2023	24/05/2023
									A070101	Failure to Capture	24/05/2023	24/05/2023
									A070102	High Capture Threshold	24/05/2023	24/05/2023
									A070104	Unstable Capture Threshold	24/05/2023	24/05/2023
							F1905	Device Revision or Replacement	A0701	Capturing Problem	24/05/2023	24/05/2023
									A070101	Failure to Capture	24/05/2023	24/05/2023
									A070102	High Capture Threshold	24/05/2023	24/05/2023
									A070104	Unstable Capture Threshold	24/05/2023	24/05/2023
					CT2750	Implantable cardiac electrical stimulators and associated devices	E060106	Heart Block	A0701	Capturing Problem	24/05/2023	24/05/2023
									A070101	Failure to Capture	24/05/2023	24/05/2023
									A070102	High Capture Threshold	24/05/2023	24/05/2023
									A070104	Unstable Capture Threshold	24/05/2023	24/05/2023
							F08	Hospitalization or Prolonged Hospitalization	A0701	Capturing Problem	24/05/2023	24/05/2023
									A070101	Failure to Capture	24/05/2023	24/05/2023
									A070102	High Capture Threshold	24/05/2023	24/05/2023
									A070104	Unstable Capture Threshold	24/05/2023	24/05/2023
							F12	Serious Injury/ Illness/ Impairment	A0701	Capturing Problem	24/05/2023	24/05/2023
									A070101	Failure to Capture	24/05/2023	24/05/2023
									A070102	High Capture Threshold	24/05/2023	24/05/2023
									A070104	Unstable Capture Threshold	24/05/2023	24/05/2023
							F1905	Device Revision or Replacement	A0701	Capturing Problem	24/05/2023	24/05/2023
									A070101	Failure to Capture	24/05/2023	24/05/2023
									A070102	High Capture Threshold	24/05/2023	24/05/2023
									A070104	Unstable Capture Threshold	24/05/2023	24/05/2023
					CT2818	Catheterization/Cannulation/Fluid path devices	E060106	Heart Block	A0701	Capturing Problem	24/05/2023	24/05/2023
									A070101	Failure to Capture	24/05/2023	24/05/2023
									A070102	High Capture Threshold	24/05/2023	24/05/2023
									A070104	Unstable Capture Threshold	24/05/2023	24/05/2023
							F08	Hospitalization or Prolonged Hospitalization	A0701	Capturing Problem	24/05/2023	24/05/2023
									A070101	Failure to Capture	24/05/2023	24/05/2023
									A070102	High Capture Threshold	24/05/2023	24/05/2023
									A070104	Unstable Capture Threshold	24/05/2023	24/05/2023
							F12	Serious Injury/ Illness/ Impairment	A0701	Capturing Problem	24/05/2023	24/05/2023
									A070101	Failure to Capture	24/05/2023	24/05/2023
									A070102	High Capture Threshold	24/05/2023	24/05/2023
									A070104	Unstable Capture Threshold	24/05/2023	24/05/2023
							F1905	Device Revision or Replacement	A0701	Capturing Problem	24/05/2023	24/05/2023
									A070101	Failure to Capture	24/05/2023	24/05/2023
									A070102	High Capture Threshold	24/05/2023	24/05/2023
									A070104	Unstable Capture Threshold	24/05/2023	24/05/2023
					CT2866	Implantable cardiac electrical stimulators	E060106	Heart Block	A0701	Capturing Problem	24/05/2023	24/05/2023
									A070101	Failure to Capture	24/05/2023	24/05/2023
									A070102	High Capture Threshold	24/05/2023	24/05/2023
									A070104	Unstable Capture Threshold	24/05/2023	24/05/2023
							F08	Hospitalization or Prolonged Hospitalization	A0701	Capturing Problem	24/05/2023	24/05/2023
									A070101	Failure to Capture	24/05/2023	24/05/2023
									A070102	High Capture Threshold	24/05/2023	24/05/2023
									A070104	Unstable Capture Threshold	24/05/2023	24/05/2023
							F12	Serious Injury/ Illness/ Impairment	A0701	Capturing Problem	24/05/2023	24/05/2023
									A070101	Failure to Capture	24/05/2023	24/05/2023
									A070102	High Capture Threshold	24/05/2023	24/05/2023
									A070104	Unstable Capture Threshold	24/05/2023	24/05/2023
							F1905	Device Revision or Replacement	A0701	Capturing Problem	24/05/2023	24/05/2023
									A070101	Failure to Capture	24/05/2023	24/05/2023
									A070102	High Capture Threshold	24/05/2023	24/05/2023
									A070104	Unstable Capture Threshold	24/05/2023	24/05/2023
					CT2985	Active	E060106	Heart Block	A0701	Capturing Problem	24/05/2023	24/05/2023
									A070101	Failure to Capture	24/05/2023	24/05/2023
									A070102	High Capture Threshold	24/05/2023	24/05/2023
									A070104	Unstable Capture Threshold	24/05/2023	24/05/2023
							F08	Hospitalization or Prolonged Hospitalization	A0701	Capturing Problem	24/05/2023	24/05/2023
									A070101	Failure to Capture	24/05/2023	24/05/2023
									A070102	High Capture Threshold	24/05/2023	24/05/2023
									A070104	Unstable Capture Threshold	24/05/2023	24/05/2023
							F12	Serious Injury/ Illness/ Impairment	A0701	Capturing Problem	24/05/2023	24/05/2023
									A070101	Failure to Capture	24/05/2023	24/05/2023
									A070102	High Capture Threshold	24/05/2023	24/05/2023
									A070104	Unstable Capture Threshold	24/05/2023	24/05/2023
							F1905	Device Revision or Replacement	A0701	Capturing Problem	24/05/2023	24/05/2023
									A070101	Failure to Capture	24/05/2023	24/05/2023
									A070102	High Capture Threshold	24/05/2023	24/05/2023
									A070104	Unstable Capture Threshold	24/05/2023	24/05/2023
					CT3001	Implantable metal	E060106	Heart Block	A0701	Capturing Problem	24/05/2023	24/05/2023
									A070101	Failure to Capture	24/05/2023	24/05/2023
									A070102	High Capture Threshold	24/05/2023	24/05/2023
									A070104	Unstable Capture Threshold	24/05/2023	24/05/2023
							F08	Hospitalization or Prolonged Hospitalization	A0701	Capturing Problem	24/05/2023	24/05/2023
									A070101	Failure to Capture	24/05/2023	24/05/2023
									A070102	High Capture Threshold	24/05/2023	24/05/2023
									A070104	Unstable Capture Threshold	24/05/2023	24/05/2023
							F12	Serious Injury/ Illness/ Impairment	A0701	Capturing Problem	24/05/2023	24/05/2023
									A070101	Failure to Capture	24/05/2023	24/05/2023
									A070102	High Capture Threshold	24/05/2023	24/05/2023
									A070104	Unstable Capture Threshold	24/05/2023	24/05/2023

						F1905	Device Revision or Replacement	A0701	Capturing Problem	24/05/2023	24/05/2023
								A070101	Failure to Capture	24/05/2023	24/05/2023
								A070102	High Capture Threshold	24/05/2023	24/05/2023
								A070104	Unstable Capture Threshold	24/05/2023	24/05/2023
					CT3002	Implantable synthetic polymer	E060106	Heart Block	A0701	Capturing Problem	24/05/2023	24/05/2023
									A070101	Failure to Capture	24/05/2023	24/05/2023
									A070102	High Capture Threshold	24/05/2023	24/05/2023
									A070104	Unstable Capture Threshold	24/05/2023	24/05/2023
							F08	Hospitalization or Prolonged Hospitalization	A0701	Capturing Problem	24/05/2023	24/05/2023
									A070101	Failure to Capture	24/05/2023	24/05/2023
									A070102	High Capture Threshold	24/05/2023	24/05/2023
									A070104	Unstable Capture Threshold	24/05/2023	24/05/2023
							F12	Serious Injury/ Illness/ Impairment	A0701	Capturing Problem	24/05/2023	24/05/2023
									A070101	Failure to Capture	24/05/2023	24/05/2023
									A070102	High Capture Threshold	24/05/2023	24/05/2023
									A070104	Unstable Capture Threshold	24/05/2023	24/05/2023
							F1905	Device Revision or Replacement	A0701	Capturing Problem	24/05/2023	24/05/2023
									A070101	Failure to Capture	24/05/2023	24/05/2023
									A070102	High Capture Threshold	24/05/2023	24/05/2023
									A070104	Unstable Capture Threshold	24/05/2023	24/05/2023
					CT3006	Implantable inorganic material	E060106	Heart Block	A0701	Capturing Problem	24/05/2023	24/05/2023
									A070101	Failure to Capture	24/05/2023	24/05/2023
									A070102	High Capture Threshold	24/05/2023	24/05/2023
									A070104	Unstable Capture Threshold	24/05/2023	24/05/2023
							F08	Hospitalization or Prolonged Hospitalization	A0701	Capturing Problem	24/05/2023	24/05/2023
									A070101	Failure to Capture	24/05/2023	24/05/2023
									A070102	High Capture Threshold	24/05/2023	24/05/2023
									A070104	Unstable Capture Threshold	24/05/2023	24/05/2023
							F12	Serious Injury/ Illness/ Impairment	A0701	Capturing Problem	24/05/2023	24/05/2023
									A070101	Failure to Capture	24/05/2023	24/05/2023
									A070102	High Capture Threshold	24/05/2023	24/05/2023
									A070104	Unstable Capture Threshold	24/05/2023	24/05/2023
							F1905	Device Revision or Replacement	A0701	Capturing Problem	24/05/2023	24/05/2023
									A070101	Failure to Capture	24/05/2023	24/05/2023
									A070102	High Capture Threshold	24/05/2023	24/05/2023
									A070104	Unstable Capture Threshold	24/05/2023	24/05/2023
					CT314	Electrically powered	E060106	Heart Block	A0701	Capturing Problem	24/05/2023	24/05/2023
									A070101	Failure to Capture	24/05/2023	24/05/2023
									A070102	High Capture Threshold	24/05/2023	24/05/2023
									A070104	Unstable Capture Threshold	24/05/2023	24/05/2023
							F08	Hospitalization or Prolonged Hospitalization	A0701	Capturing Problem	24/05/2023	24/05/2023
									A070101	Failure to Capture	24/05/2023	24/05/2023
									A070102	High Capture Threshold	24/05/2023	24/05/2023
									A070104	Unstable Capture Threshold	24/05/2023	24/05/2023
							F12	Serious Injury/ Illness/ Impairment	A0701	Capturing Problem	24/05/2023	24/05/2023
									A070101	Failure to Capture	24/05/2023	24/05/2023
									A070102	High Capture Threshold	24/05/2023	24/05/2023
									A070104	Unstable Capture Threshold	24/05/2023	24/05/2023
							F1905	Device Revision or Replacement	A0701	Capturing Problem	24/05/2023	24/05/2023
									A070101	Failure to Capture	24/05/2023	24/05/2023
									A070102	High Capture Threshold	24/05/2023	24/05/2023
									A070104	Unstable Capture Threshold	24/05/2023	24/05/2023
					CT321	Long-term surgical invasive	E060106	Heart Block	A0701	Capturing Problem	24/05/2023	24/05/2023
									A070101	Failure to Capture	24/05/2023	24/05/2023
									A070102	High Capture Threshold	24/05/2023	24/05/2023
									A070104	Unstable Capture Threshold	24/05/2023	24/05/2023
							F08	Hospitalization or Prolonged Hospitalization	A0701	Capturing Problem	24/05/2023	24/05/2023
									A070101	Failure to Capture	24/05/2023	24/05/2023
									A070102	High Capture Threshold	24/05/2023	24/05/2023
									A070104	Unstable Capture Threshold	24/05/2023	24/05/2023
							F12	Serious Injury/ Illness/ Impairment	A0701	Capturing Problem	24/05/2023	24/05/2023
									A070101	Failure to Capture	24/05/2023	24/05/2023
									A070102	High Capture Threshold	24/05/2023	24/05/2023
									A070104	Unstable Capture Threshold	24/05/2023	24/05/2023
							F1905	Device Revision or Replacement	A0701	Capturing Problem	24/05/2023	24/05/2023
									A070101	Failure to Capture	24/05/2023	24/05/2023
									A070102	High Capture Threshold	24/05/2023	24/05/2023
									A070104	Unstable Capture Threshold	24/05/2023	24/05/2023
					CT981	Single use	E060106	Heart Block	A0701	Capturing Problem	24/05/2023	24/05/2023
									A070101	Failure to Capture	24/05/2023	24/05/2023
									A070102	High Capture Threshold	24/05/2023	24/05/2023
									A070104	Unstable Capture Threshold	24/05/2023	24/05/2023
							F08	Hospitalization or Prolonged Hospitalization	A0701	Capturing Problem	24/05/2023	24/05/2023
									A070101	Failure to Capture	24/05/2023	24/05/2023
									A070102	High Capture Threshold	24/05/2023	24/05/2023
									A070104	Unstable Capture Threshold	24/05/2023	24/05/2023
							F12	Serious Injury/ Illness/ Impairment	A0701	Capturing Problem	24/05/2023	24/05/2023
									A070101	Failure to Capture	24/05/2023	24/05/2023
									A070102	High Capture Threshold	24/05/2023	24/05/2023
									A070104	Unstable Capture Threshold	24/05/2023	24/05/2023
							F1905	Device Revision or Replacement	A0701	Capturing Problem	24/05/2023	24/05/2023
									A070101	Failure to Capture	24/05/2023	24/05/2023
									A070102	High Capture Threshold	24/05/2023	24/05/2023
									A070104	Unstable Capture Threshold	24/05/2023	24/05/2023
					CT983	Surgical invasive	E060106	Heart Block	A0701	Capturing Problem	24/05/2023	24/05/2023
									A070101	Failure to Capture	24/05/2023	24/05/2023
									A070102	High Capture Threshold	24/05/2023	24/05/2023
									A070104	Unstable Capture Threshold	24/05/2023	24/05/2023

							F08	Hospitalization or Prolonged Hospitalization	A0701	Capturing Problem	24/05/2023	24/05/2023
									A070101	Failure to Capture	24/05/2023	24/05/2023
									A070102	High Capture Threshold	24/05/2023	24/05/2023
									A070104	Unstable Capture Threshold	24/05/2023	24/05/2023
							F12	Serious Injury/ Illness/ Impairment	A0701	Capturing Problem	24/05/2023	24/05/2023
									A070101	Failure to Capture	24/05/2023	24/05/2023
									A070102	High Capture Threshold	24/05/2023	24/05/2023
									A070104	Unstable Capture Threshold	24/05/2023	24/05/2023
							F1905	Device Revision or Replacement	A0701	Capturing Problem	24/05/2023	24/05/2023
									A070101	Failure to Capture	24/05/2023	24/05/2023
									A070102	High Capture Threshold	24/05/2023	24/05/2023
									A070104	Unstable Capture Threshold	24/05/2023	24/05/2023
13832	2023	All other reportable incidents	47267	Single-chamber implantable pacemaker, rate-responsive	CT141	Electrodes/Leads and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT145	Cardiovascular devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT233	Surgical	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT2406	Implantable	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT2407	Active implantable	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT2727	Active implant devices/ancillaries	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT2750	Implantable cardiac electrical stimulators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT2985	Active	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT3001	Implantable metal	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT3002	Implantable synthetic polymer	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT3006	Implantable inorganic material	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT314	Electrically powered	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT321	Long-term surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT449	Generators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT981	Single use	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT983	Surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
13833	2023	All other reportable incidents	36241	Pacemaker/defibrillator lead	CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A072201	High Impedance		
							F26	No Health Consequences or Impact	A072201	High Impedance		
13834	2023	All other reportable incidents	47265	Dual-chamber implantable pacemaker, rate-responsive	CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A05	Mechanical Problem	10/08/2023	10/08/2023
							F08	Hospitalization or Prolonged Hospitalization	A05	Mechanical Problem	10/08/2023	10/08/2023
							F12	Serious Injury/ Illness/ Impairment	A05	Mechanical Problem	10/08/2023	10/08/2023
							F1905	Device Revision or Replacement	A05	Mechanical Problem	10/08/2023	10/08/2023
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A05	Mechanical Problem	10/08/2023	10/08/2023
							F08	Hospitalization or Prolonged Hospitalization	A05	Mechanical Problem	10/08/2023	10/08/2023
							F12	Serious Injury/ Illness/ Impairment	A05	Mechanical Problem	10/08/2023	10/08/2023
							F1905	Device Revision or Replacement	A05	Mechanical Problem	10/08/2023	10/08/2023
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A05	Mechanical Problem	10/08/2023	10/08/2023
							F08	Hospitalization or Prolonged Hospitalization	A05	Mechanical Problem	10/08/2023	10/08/2023
							F12	Serious Injury/ Illness/ Impairment	A05	Mechanical Problem	10/08/2023	10/08/2023
							F1905	Device Revision or Replacement	A05	Mechanical Problem	10/08/2023	10/08/2023
13835	2023	All other reportable incidents	35223	Endocardial/interventricular septal pacing lead	CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A072201	High Impedance		
							F1904	Device Repositioning	A072201	High Impedance		
							F1905	Device Revision or Replacement	A072201	High Impedance		
13836	2023	All other reportable incidents	35822	External pacemaker, transcutaneous pacing	CT1684	Cardiac pacemakers and associated devices	E0602	Cardiac Arrest	A070503	Failure to Run on Battery		
							F2306	Resuscitation	A070503	Failure to Run on Battery		
					CT1685	Cardiac pacemakers	E0602	Cardiac Arrest	A070503	Failure to Run on Battery		
							F2306	Resuscitation	A070503	Failure to Run on Battery		
13837	2023	All other reportable incidents	35223	Endocardial/interventricular septal pacing lead	CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070102	High Capture Threshold		
									A070901	Decreased Sensitivity		
									A070909	Oversensing		
									A090801	Signal Artifact/Noise		
							F12	Serious Injury/ Illness/ Impairment	A070102	High Capture Threshold		
									A070901	Decreased Sensitivity		
									A070909	Oversensing		
									A090801	Signal Artifact/Noise		
							F23	Unexpected Medical Intervention	A070102	High Capture Threshold		
									A070901	Decreased Sensitivity		
									A070909	Oversensing		
									A090801	Signal Artifact/Noise		
13838	2023	All other reportable incidents	35223	Endocardial/interventricular septal pacing lead	CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A040101	Fracture	18/04/2023	18/04/2023
									A070909	Oversensing	18/04/2023	18/04/2023
							F1901	Additional Surgery	A040101	Fracture	18/04/2023	18/04/2023

									A070909	Oversensing	18/04/2023	18/04/2023
13839	2023	All other reportable incidents	47265	Dual-chamber implantable pacemaker, rate-responsive	CT141	Electrodes/Leads and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A090804	False Positive Result		
							F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A090804	False Positive Result		
					CT145	Cardiovascular devices	E2403	No Clinical Signs, Symptoms or Conditions	A090804	False Positive Result		
							F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A090804	False Positive Result		
					CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A090804	False Positive Result		
							F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A090804	False Positive Result		
					CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A090804	False Positive Result		
							F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A090804	False Positive Result		
					CT233	Surgical	E2403	No Clinical Signs, Symptoms or Conditions	A090804	False Positive Result		
							F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A090804	False Positive Result		
					CT2406	Implantable	E2403	No Clinical Signs, Symptoms or Conditions	A090804	False Positive Result		
							F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A090804	False Positive Result		
					CT2407	Active implantable	E2403	No Clinical Signs, Symptoms or Conditions	A090804	False Positive Result		
							F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A090804	False Positive Result		
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A090804	False Positive Result		
							F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A090804	False Positive Result		
					CT2727	Active implant devices/ancillaries	E2403	No Clinical Signs, Symptoms or Conditions	A090804	False Positive Result		
							F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A090804	False Positive Result		
					CT2750	Implantable cardiac electrical stimulators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A090804	False Positive Result		
							F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A090804	False Positive Result		
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A090804	False Positive Result		
							F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A090804	False Positive Result		
					CT2985	Active	E2403	No Clinical Signs, Symptoms or Conditions	A090804	False Positive Result		
							F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A090804	False Positive Result		
					CT3001	Implantable metal	E2403	No Clinical Signs, Symptoms or Conditions	A090804	False Positive Result		
							F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A090804	False Positive Result		
					CT3002	Implantable synthetic polymer	E2403	No Clinical Signs, Symptoms or Conditions	A090804	False Positive Result		
							F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A090804	False Positive Result		
					CT3006	Implantable inorganic material	E2403	No Clinical Signs, Symptoms or Conditions	A090804	False Positive Result		
							F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A090804	False Positive Result		
					CT314	Electrically powered	E2403	No Clinical Signs, Symptoms or Conditions	A090804	False Positive Result		
							F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A090804	False Positive Result		
					CT321	Long-term surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A090804	False Positive Result		
							F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A090804	False Positive Result		
					CT449	Generators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A090804	False Positive Result		
							F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A090804	False Positive Result		
					CT981	Single use	E2403	No Clinical Signs, Symptoms or Conditions	A090804	False Positive Result		
							F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A090804	False Positive Result		
					CT983	Surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A090804	False Positive Result		
							F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A090804	False Positive Result		
13840	2023	All other reportable incidents	47265	Dual-chamber implantable pacemaker, rate-responsive	CT141	Electrodes/Leads and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem	02/10/2023	02/10/2023
									A070504	Premature Discharge of Battery	02/10/2023	02/10/2023
							F08	Hospitalization or Prolonged Hospitalization	A0705	Battery Problem	02/10/2023	02/10/2023
									A070504	Premature Discharge of Battery	02/10/2023	02/10/2023
							F1905	Device Revision or Replacement	A0705	Battery Problem	02/10/2023	02/10/2023
									A070504	Premature Discharge of Battery	02/10/2023	02/10/2023
					CT233	Surgical	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem	02/10/2023	02/10/2023
									A070504	Premature Discharge of Battery	02/10/2023	02/10/2023
							F08	Hospitalization or Prolonged Hospitalization	A0705	Battery Problem	02/10/2023	02/10/2023
									A070504	Premature Discharge of Battery	02/10/2023	02/10/2023
							F1905	Device Revision or Replacement	A0705	Battery Problem	02/10/2023	02/10/2023
									A070504	Premature Discharge of Battery	02/10/2023	02/10/2023
					CT2407	Active implantable	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem	02/10/2023	02/10/2023
									A070504	Premature Discharge of Battery	02/10/2023	02/10/2023
							F08	Hospitalization or Prolonged Hospitalization	A0705	Battery Problem	02/10/2023	02/10/2023
									A070504	Premature Discharge of Battery	02/10/2023	02/10/2023
							F1905	Device Revision or Replacement	A0705	Battery Problem	02/10/2023	02/10/2023
									A070504	Premature Discharge of Battery	02/10/2023	02/10/2023
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem	02/10/2023	02/10/2023
									A070504	Premature Discharge of Battery	02/10/2023	02/10/2023
							F08	Hospitalization or Prolonged Hospitalization	A0705	Battery Problem	02/10/2023	02/10/2023
									A070504	Premature Discharge of Battery	02/10/2023	02/10/2023
							F1905	Device Revision or Replacement	A0705	Battery Problem	02/10/2023	02/10/2023
									A070504	Premature Discharge of Battery	02/10/2023	02/10/2023
					CT3001	Implantable metal	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem	02/10/2023	02/10/2023
									A070504	Premature Discharge of Battery	02/10/2023	02/10/2023
							F08	Hospitalization or Prolonged Hospitalization	A0705	Battery Problem	02/10/2023	02/10/2023
									A070504	Premature Discharge of Battery	02/10/2023	02/10/2023
							F1905	Device Revision or Replacement	A0705	Battery Problem	02/10/2023	02/10/2023
									A070504	Premature Discharge of Battery	02/10/2023	02/10/2023
					CT3002	Implantable synthetic polymer	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem	02/10/2023	02/10/2023
									A070504	Premature Discharge of Battery	02/10/2023	02/10/2023
							F08	Hospitalization or Prolonged Hospitalization	A0705	Battery Problem	02/10/2023	02/10/2023
									A070504	Premature Discharge of Battery	02/10/2023	02/10/2023
							F1905	Device Revision or Replacement	A0705	Battery Problem	02/10/2023	02/10/2023
									A070504	Premature Discharge of Battery	02/10/2023	02/10/2023
					CT314	Electrically powered	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem	02/10/2023	02/10/2023
									A070504	Premature Discharge of Battery	02/10/2023	02/10/2023
							F08	Hospitalization or Prolonged Hospitalization	A0705	Battery Problem	02/10/2023	02/10/2023
									A070504	Premature Discharge of Battery	02/10/2023	02/10/2023
							F1905	Device Revision or Replacement	A0705	Battery Problem	02/10/2023	02/10/2023
									A070504	Premature Discharge of Battery	02/10/2023	02/10/2023
					CT321	Long-term surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem	02/10/2023	02/10/2023
									A070504	Premature Discharge of Battery	02/10/2023	02/10/2023
							F08	Hospitalization or Prolonged Hospitalization	A0705	Battery Problem	02/10/2023	02/10/2023
									A070504	Premature Discharge of Battery	02/10/2023	02/10/2023
							F1905	Device Revision or Replacement	A0705	Battery Problem	02/10/2023	02/10/2023

									A070504	Premature Discharge of Battery	02/10/2023	02/10/2023
					CT449	Generators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem	02/10/2023	02/10/2023
									A070504	Premature Discharge of Battery	02/10/2023	02/10/2023
							F08	Hospitalization or Prolonged Hospitalization	A0705	Battery Problem	02/10/2023	02/10/2023
									A070504	Premature Discharge of Battery	02/10/2023	02/10/2023
							F1905	Device Revision or Replacement	A0705	Battery Problem	02/10/2023	02/10/2023
									A070504	Premature Discharge of Battery	02/10/2023	02/10/2023
					CT981	Single use	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem	02/10/2023	02/10/2023
									A070504	Premature Discharge of Battery	02/10/2023	02/10/2023
							F08	Hospitalization or Prolonged Hospitalization	A0705	Battery Problem	02/10/2023	02/10/2023
									A070504	Premature Discharge of Battery	02/10/2023	02/10/2023
							F1905	Device Revision or Replacement	A0705	Battery Problem	02/10/2023	02/10/2023
									A070504	Premature Discharge of Battery	02/10/2023	02/10/2023
							E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
							F24	Insufficient Health Impact Information	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
13841	2023	All other reportable incidents	47263	Cardiac resynchronization therapy implantable pacemaker	CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A090804	False Positive Result		
							F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A090804	False Positive Result		
							F15	Recognized Device or Procedural Complication	A090804	False Positive Result		
					CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A090804	False Positive Result		
							F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A090804	False Positive Result		
							F15	Recognized Device or Procedural Complication	A090804	False Positive Result		
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A090804	False Positive Result		
							F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A090804	False Positive Result		
							F15	Recognized Device or Procedural Complication	A090804	False Positive Result		
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A090804	False Positive Result		
							F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A090804	False Positive Result		
							F15	Recognized Device or Procedural Complication	A090804	False Positive Result		
13842	2023	All other reportable incidents	17882	Physiologic-monitoring defibrillation system	CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070603	Failure to Charge		
							F27	Problem Identified During Non-Clinical Procedure	A070603	Failure to Charge		
13843	2023	All other reportable incidents	35223	Endocardial/interventricular septal pacing lead	CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A0401	Break		
							F08	Hospitalization or Prolonged Hospitalization	A0401	Break		
							F12	Serious Injury/ Illness/ Impairment	A0401	Break		
							F1905	Device Revision or Replacement	A0401	Break		
13844	2023	All other reportable incidents	35223	Endocardial/interventricular septal pacing lead	CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070102	High Capture Threshold		
							F12	Serious Injury/ Illness/ Impairment	A070102	High Capture Threshold		
							F23	Unexpected Medical Intervention	A070102	High Capture Threshold		
13845	2023	All other reportable incidents	47265	Dual-chamber implantable pacemaker, rate-responsive	CT141	Electrodes/Leads and associated devices	E011903	Syncope/Fainting	A05	Mechanical Problem	16/08/2023	16/08/2023
							E060101	Asystole	A05	Mechanical Problem	16/08/2023	16/08/2023
							F08	Hospitalization or Prolonged Hospitalization	A05	Mechanical Problem	16/08/2023	16/08/2023
							F1203	Life Threatening Illness or Injury	A05	Mechanical Problem	16/08/2023	16/08/2023
							F1905	Device Revision or Replacement	A05	Mechanical Problem	16/08/2023	16/08/2023
					CT145	Cardiovascular devices	E011903	Syncope/Fainting	A05	Mechanical Problem	16/08/2023	16/08/2023
							E060101	Asystole	A05	Mechanical Problem	16/08/2023	16/08/2023
							F08	Hospitalization or Prolonged Hospitalization	A05	Mechanical Problem	16/08/2023	16/08/2023
							F1203	Life Threatening Illness or Injury	A05	Mechanical Problem	16/08/2023	16/08/2023
							F1905	Device Revision or Replacement	A05	Mechanical Problem	16/08/2023	16/08/2023
					CT1684	Cardiac pacemakers and associated devices	E011903	Syncope/Fainting	A05	Mechanical Problem	16/08/2023	16/08/2023
							E060101	Asystole	A05	Mechanical Problem	16/08/2023	16/08/2023
							F08	Hospitalization or Prolonged Hospitalization	A05	Mechanical Problem	16/08/2023	16/08/2023
							F1203	Life Threatening Illness or Injury	A05	Mechanical Problem	16/08/2023	16/08/2023
							F1905	Device Revision or Replacement	A05	Mechanical Problem	16/08/2023	16/08/2023
					CT1685	Cardiac pacemakers	E011903	Syncope/Fainting	A05	Mechanical Problem	16/08/2023	16/08/2023
							E060101	Asystole	A05	Mechanical Problem	16/08/2023	16/08/2023
							F08	Hospitalization or Prolonged Hospitalization	A05	Mechanical Problem	16/08/2023	16/08/2023
							F1203	Life Threatening Illness or Injury	A05	Mechanical Problem	16/08/2023	16/08/2023
							F1905	Device Revision or Replacement	A05	Mechanical Problem	16/08/2023	16/08/2023
					CT233	Surgical	E011903	Syncope/Fainting	A05	Mechanical Problem	16/08/2023	16/08/2023
							E060101	Asystole	A05	Mechanical Problem	16/08/2023	16/08/2023
							F08	Hospitalization or Prolonged Hospitalization	A05	Mechanical Problem	16/08/2023	16/08/2023
							F1203	Life Threatening Illness or Injury	A05	Mechanical Problem	16/08/2023	16/08/2023
							F1905	Device Revision or Replacement	A05	Mechanical Problem	16/08/2023	16/08/2023
					CT2406	Implantable	E011903	Syncope/Fainting	A05	Mechanical Problem	16/08/2023	16/08/2023
							E060101	Asystole	A05	Mechanical Problem	16/08/2023	16/08/2023
							F08	Hospitalization or Prolonged Hospitalization	A05	Mechanical Problem	16/08/2023	16/08/2023
							F1203	Life Threatening Illness or Injury	A05	Mechanical Problem	16/08/2023	16/08/2023
							F1905	Device Revision or Replacement	A05	Mechanical Problem	16/08/2023	16/08/2023
					CT2407	Active implantable	E011903	Syncope/Fainting	A05	Mechanical Problem	16/08/2023	16/08/2023
							E060101	Asystole	A05	Mechanical Problem	16/08/2023	16/08/2023
							F08	Hospitalization or Prolonged Hospitalization	A05	Mechanical Problem	16/08/2023	16/08/2023
							F1203	Life Threatening Illness or Injury	A05	Mechanical Problem	16/08/2023	16/08/2023
							F1905	Device Revision or Replacement	A05	Mechanical Problem	16/08/2023	16/08/2023
					CT257	Implantable cardiac pacemakers	E011903	Syncope/Fainting	A05	Mechanical Problem	16/08/2023	16/08/2023
							E060101	Asystole	A05	Mechanical Problem	16/08/2023	16/08/2023
							F08	Hospitalization or Prolonged Hospitalization	A05	Mechanical Problem	16/08/2023	16/08/2023
							F1203	Life Threatening Illness or Injury	A05	Mechanical Problem	16/08/2023	16/08/2023
							F1905	Device Revision or Replacement	A05	Mechanical Problem	16/08/2023	16/08/2023
					CT2727	Active implant devices/ancillaries	E011903	Syncope/Fainting	A05	Mechanical Problem	16/08/2023	16/08/2023
							E060101	Asystole	A05	Mechanical Problem	16/08/2023	16/08/2023
							F08	Hospitalization or Prolonged Hospitalization	A05	Mechanical Problem	16/08/2023	16/08/2023
							F1203	Life Threatening Illness or Injury	A05	Mechanical Problem	16/08/2023	16/08/2023
							F1905	Device Revision or Replacement	A05	Mechanical Problem	16/08/2023	16/08/2023
					CT2750	Implantable cardiac electrical stimulators and associated devices	E011903	Syncope/Fainting	A05	Mechanical Problem	16/08/2023	16/08/2023
							E060101	Asystole	A05	Mechanical Problem	16/08/2023	16/08/2023
							F08	Hospitalization or Prolonged Hospitalization	A05	Mechanical Problem	16/08/2023	16/08/2023
							F1203	Life Threatening Illness or Injury	A05	Mechanical Problem	16/08/2023	16/08/2023
							F1905	Device Revision or Replacement	A05	Mechanical Problem	16/08/2023	16/08/2023

					CT2866	Implantable cardiac electrical stimulators	E011903	Syncope/Fainting	A05	Mechanical Problem	16/08/2023	16/08/2023
							E060101	Asystole	A05	Mechanical Problem	16/08/2023	16/08/2023
							F08	Hospitalization or Prolonged Hospitalization	A05	Mechanical Problem	16/08/2023	16/08/2023
							F1203	Life Threatening Illness or Injury	A05	Mechanical Problem	16/08/2023	16/08/2023
							F1905	Device Revision or Replacement	A05	Mechanical Problem	16/08/2023	16/08/2023
					CT2985	Active	E011903	Syncope/Fainting	A05	Mechanical Problem	16/08/2023	16/08/2023
							E060101	Asystole	A05	Mechanical Problem	16/08/2023	16/08/2023
							F08	Hospitalization or Prolonged Hospitalization	A05	Mechanical Problem	16/08/2023	16/08/2023
							F1203	Life Threatening Illness or Injury	A05	Mechanical Problem	16/08/2023	16/08/2023
							F1905	Device Revision or Replacement	A05	Mechanical Problem	16/08/2023	16/08/2023
					CT3001	Implantable metal	E011903	Syncope/Fainting	A05	Mechanical Problem	16/08/2023	16/08/2023
							E060101	Asystole	A05	Mechanical Problem	16/08/2023	16/08/2023
							F08	Hospitalization or Prolonged Hospitalization	A05	Mechanical Problem	16/08/2023	16/08/2023
							F1203	Life Threatening Illness or Injury	A05	Mechanical Problem	16/08/2023	16/08/2023
							F1905	Device Revision or Replacement	A05	Mechanical Problem	16/08/2023	16/08/2023
					CT3002	Implantable synthetic polymer	E011903	Syncope/Fainting	A05	Mechanical Problem	16/08/2023	16/08/2023
							E060101	Asystole	A05	Mechanical Problem	16/08/2023	16/08/2023
							F08	Hospitalization or Prolonged Hospitalization	A05	Mechanical Problem	16/08/2023	16/08/2023
							F1203	Life Threatening Illness or Injury	A05	Mechanical Problem	16/08/2023	16/08/2023
							F1905	Device Revision or Replacement	A05	Mechanical Problem	16/08/2023	16/08/2023
					CT3006	Implantable inorganic material	E011903	Syncope/Fainting	A05	Mechanical Problem	16/08/2023	16/08/2023
							E060101	Asystole	A05	Mechanical Problem	16/08/2023	16/08/2023
							F08	Hospitalization or Prolonged Hospitalization	A05	Mechanical Problem	16/08/2023	16/08/2023
							F1203	Life Threatening Illness or Injury	A05	Mechanical Problem	16/08/2023	16/08/2023
							F1905	Device Revision or Replacement	A05	Mechanical Problem	16/08/2023	16/08/2023
					CT314	Electrically powered	E011903	Syncope/Fainting	A05	Mechanical Problem	16/08/2023	16/08/2023
							E060101	Asystole	A05	Mechanical Problem	16/08/2023	16/08/2023
							F08	Hospitalization or Prolonged Hospitalization	A05	Mechanical Problem	16/08/2023	16/08/2023
							F1203	Life Threatening Illness or Injury	A05	Mechanical Problem	16/08/2023	16/08/2023
							F1905	Device Revision or Replacement	A05	Mechanical Problem	16/08/2023	16/08/2023
					CT321	Long-term surgical invasive	E011903	Syncope/Fainting	A05	Mechanical Problem	16/08/2023	16/08/2023
							E060101	Asystole	A05	Mechanical Problem	16/08/2023	16/08/2023
							F08	Hospitalization or Prolonged Hospitalization	A05	Mechanical Problem	16/08/2023	16/08/2023
							F1203	Life Threatening Illness or Injury	A05	Mechanical Problem	16/08/2023	16/08/2023
							F1905	Device Revision or Replacement	A05	Mechanical Problem	16/08/2023	16/08/2023
					CT449	Generators and associated devices	E011903	Syncope/Fainting	A05	Mechanical Problem	16/08/2023	16/08/2023
							E060101	Asystole	A05	Mechanical Problem	16/08/2023	16/08/2023
							F08	Hospitalization or Prolonged Hospitalization	A05	Mechanical Problem	16/08/2023	16/08/2023
							F1203	Life Threatening Illness or Injury	A05	Mechanical Problem	16/08/2023	16/08/2023
							F1905	Device Revision or Replacement	A05	Mechanical Problem	16/08/2023	16/08/2023
					CT981	Single use	E011903	Syncope/Fainting	A05	Mechanical Problem	16/08/2023	16/08/2023
							E060101	Asystole	A05	Mechanical Problem	16/08/2023	16/08/2023
							F08	Hospitalization or Prolonged Hospitalization	A05	Mechanical Problem	16/08/2023	16/08/2023
							F1203	Life Threatening Illness or Injury	A05	Mechanical Problem	16/08/2023	16/08/2023
							F1905	Device Revision or Replacement	A05	Mechanical Problem	16/08/2023	16/08/2023
					CT983	Surgical invasive	E011903	Syncope/Fainting	A05	Mechanical Problem	16/08/2023	16/08/2023
							E060101	Asystole	A05	Mechanical Problem	16/08/2023	16/08/2023
							F08	Hospitalization or Prolonged Hospitalization	A05	Mechanical Problem	16/08/2023	16/08/2023
							F1203	Life Threatening Illness or Injury	A05	Mechanical Problem	16/08/2023	16/08/2023
							F1905	Device Revision or Replacement	A05	Mechanical Problem	16/08/2023	16/08/2023
13846	2023	All other reportable incidents	47263	Cardiac resynchronization therapy implantable pacemaker	CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A0908	Incorrect, Inadequate or Imprecise Result or Readings	08/03/2023	08/03/2023
							F08	Hospitalization or Prolonged Hospitalization	A0908	Incorrect, Inadequate or Imprecise Result or Readings	08/03/2023	08/03/2023
							F1905	Device Revision or Replacement	A0908	Incorrect, Inadequate or Imprecise Result or Readings	08/03/2023	08/03/2023
					CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A0908	Incorrect, Inadequate or Imprecise Result or Readings	08/03/2023	08/03/2023
							F08	Hospitalization or Prolonged Hospitalization	A0908	Incorrect, Inadequate or Imprecise Result or Readings	08/03/2023	08/03/2023
							F1905	Device Revision or Replacement	A0908	Incorrect, Inadequate or Imprecise Result or Readings	08/03/2023	08/03/2023
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A0908	Incorrect, Inadequate or Imprecise Result or Readings	08/03/2023	08/03/2023
							F08	Hospitalization or Prolonged Hospitalization	A0908	Incorrect, Inadequate or Imprecise Result or Readings	08/03/2023	08/03/2023
							F1905	Device Revision or Replacement	A0908	Incorrect, Inadequate or Imprecise Result or Readings	08/03/2023	08/03/2023
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A0908	Incorrect, Inadequate or Imprecise Result or Readings	08/03/2023	08/03/2023
							F08	Hospitalization or Prolonged Hospitalization	A0908	Incorrect, Inadequate or Imprecise Result or Readings	08/03/2023	08/03/2023
							F1905	Device Revision or Replacement	A0908	Incorrect, Inadequate or Imprecise Result or Readings	08/03/2023	08/03/2023
13847	2023	All other reportable incidents	47265	Dual-chamber implantable pacemaker, rate-responsive	CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem	08/06/2023	08/06/2023
									A070504	Premature Discharge of Battery	08/06/2023	08/06/2023
							F08	Hospitalization or Prolonged Hospitalization	A0705	Battery Problem	08/06/2023	08/06/2023
									A070504	Premature Discharge of Battery	08/06/2023	08/06/2023
							F1905	Device Revision or Replacement	A0705	Battery Problem	08/06/2023	08/06/2023
									A070504	Premature Discharge of Battery	08/06/2023	08/06/2023
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem	08/06/2023	08/06/2023
									A070504	Premature Discharge of Battery	08/06/2023	08/06/2023
							F08	Hospitalization or Prolonged Hospitalization	A0705	Battery Problem	08/06/2023	08/06/2023
									A070504	Premature Discharge of Battery	08/06/2023	08/06/2023
							F1905	Device Revision or Replacement	A0705	Battery Problem	08/06/2023	08/06/2023
									A070504	Premature Discharge of Battery	08/06/2023	08/06/2023
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem	08/06/2023	08/06/2023
									A070504	Premature Discharge of Battery	08/06/2023	08/06/2023
							F08	Hospitalization or Prolonged Hospitalization	A0705	Battery Problem	08/06/2023	08/06/2023
									A070504	Premature Discharge of Battery	08/06/2023	08/06/2023
							F1905	Device Revision or Replacement	A0705	Battery Problem	08/06/2023	08/06/2023

									A070504	Premature Discharge of Battery	08/06/2023	08/06/2023
13848	2023	All other reportable incidents	47265	Dual-chamber implantable pacemaker, rate-responsive	CT141	Electrodes/Leads and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
					CT145	Cardiovascular devices	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
					CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
					CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
					CT233	Surgical	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
					CT2406	Implantable	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
					CT2407	Active implantable	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
					CT2727	Active implant devices/ancillaries	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
					CT2750	Implantable cardiac electrical stimulators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
					CT2985	Active	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
					CT3001	Implantable metal	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
					CT3002	Implantable synthetic polymer	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
					CT3006	Implantable inorganic material	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
					CT314	Electrically powered	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
					CT321	Long-term surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
					CT449	Generators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
					CT981	Single use	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
					CT983	Surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
13849	2023	All other reportable incidents	36241	Pacemaker/defibrillator lead	CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A0203	Defective Device	02/08/2023	02/08/2023
									A070102	High Capture Threshold	02/08/2023	02/08/2023
							F26	No Health Consequences or Impact	A0203	Defective Device	02/08/2023	02/08/2023
									A070102	High Capture Threshold	02/08/2023	02/08/2023
13850	2023	All other reportable incidents	35852	Single-chamber implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	22/09/2023	22/09/2023
									A0908	Incorrect, Inadequate or Imprecise Result or Readings	22/09/2023	22/09/2023

							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	22/09/2023	22/09/2023
									A0908	Incorrect, Inadequate or Imprecise Result or Readings	22/09/2023	22/09/2023
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	22/09/2023	22/09/2023
									A0908	Incorrect, Inadequate or Imprecise Result or Readings	22/09/2023	22/09/2023
13851	2023	All other reportable incidents	47263	Cardiac resynchronization therapy implantable pacemaker	CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A0908	Incorrect, Inadequate or Imprecise Result or Readings		
							F26	No Health Consequences or Impact	A0908	Incorrect, Inadequate or Imprecise Result or Readings		
					CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A0908	Incorrect, Inadequate or Imprecise Result or Readings		
							F26	No Health Consequences or Impact	A0908	Incorrect, Inadequate or Imprecise Result or Readings		
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A0908	Incorrect, Inadequate or Imprecise Result or Readings		
							F26	No Health Consequences or Impact	A0908	Incorrect, Inadequate or Imprecise Result or Readings		
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A0908	Incorrect, Inadequate or Imprecise Result or Readings		
							F26	No Health Consequences or Impact	A0908	Incorrect, Inadequate or Imprecise Result or Readings		
13852	2023	All other reportable incidents	36241	Pacemaker/defibrillator lead	CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A090801	Signal Artifact/Noise	19/04/2023	19/04/2023
							F1908	Prolonged Surgery	A090801	Signal Artifact/Noise	19/04/2023	19/04/2023
13853	2023	All other reportable incidents	35852	Single-chamber implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	03/11/2023	03/11/2023
									A0908	Incorrect, Inadequate or Imprecise Result or Readings	03/11/2023	03/11/2023
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	03/11/2023	03/11/2023
									A0908	Incorrect, Inadequate or Imprecise Result or Readings	03/11/2023	03/11/2023
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	03/11/2023	03/11/2023
									A0908	Incorrect, Inadequate or Imprecise Result or Readings	03/11/2023	03/11/2023
13854	2023	All other reportable incidents	47265	Dual-chamber implantable pacemaker, rate-responsive	CT141	Electrodes/Leads and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A0908	Incorrect, Inadequate or Imprecise Result or Readings	04/08/2023	04/08/2023
							F08	Hospitalization or Prolonged Hospitalization	A0908	Incorrect, Inadequate or Imprecise Result or Readings	04/08/2023	04/08/2023
							F1905	Device Revision or Replacement	A0908	Incorrect, Inadequate or Imprecise Result or Readings	04/08/2023	04/08/2023
					CT145	Cardiovascular devices	E2403	No Clinical Signs, Symptoms or Conditions	A0908	Incorrect, Inadequate or Imprecise Result or Readings	04/08/2023	04/08/2023
							F08	Hospitalization or Prolonged Hospitalization	A0908	Incorrect, Inadequate or Imprecise Result or Readings	04/08/2023	04/08/2023
							F1905	Device Revision or Replacement	A0908	Incorrect, Inadequate or Imprecise Result or Readings	04/08/2023	04/08/2023
					CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A0908	Incorrect, Inadequate or Imprecise Result or Readings	04/08/2023	04/08/2023
							F08	Hospitalization or Prolonged Hospitalization	A0908	Incorrect, Inadequate or Imprecise Result or Readings	04/08/2023	04/08/2023
							F1905	Device Revision or Replacement	A0908	Incorrect, Inadequate or Imprecise Result or Readings	04/08/2023	04/08/2023
					CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A0908	Incorrect, Inadequate or Imprecise Result or Readings	04/08/2023	04/08/2023
							F08	Hospitalization or Prolonged Hospitalization	A0908	Incorrect, Inadequate or Imprecise Result or Readings	04/08/2023	04/08/2023
							F1905	Device Revision or Replacement	A0908	Incorrect, Inadequate or Imprecise Result or Readings	04/08/2023	04/08/2023
					CT233	Surgical	E2403	No Clinical Signs, Symptoms or Conditions	A0908	Incorrect, Inadequate or Imprecise Result or Readings	04/08/2023	04/08/2023
							F08	Hospitalization or Prolonged Hospitalization	A0908	Incorrect, Inadequate or Imprecise Result or Readings	04/08/2023	04/08/2023
							F1905	Device Revision or Replacement	A0908	Incorrect, Inadequate or Imprecise Result or Readings	04/08/2023	04/08/2023
					CT2406	Implantable	E2403	No Clinical Signs, Symptoms or Conditions	A0908	Incorrect, Inadequate or Imprecise Result or Readings	04/08/2023	04/08/2023
							F08	Hospitalization or Prolonged Hospitalization	A0908	Incorrect, Inadequate or Imprecise Result or Readings	04/08/2023	04/08/2023
							F1905	Device Revision or Replacement	A0908	Incorrect, Inadequate or Imprecise Result or Readings	04/08/2023	04/08/2023
					CT2407	Active implantable	E2403	No Clinical Signs, Symptoms or Conditions	A0908	Incorrect, Inadequate or Imprecise Result or Readings	04/08/2023	04/08/2023
							F08	Hospitalization or Prolonged Hospitalization	A0908	Incorrect, Inadequate or Imprecise Result or Readings	04/08/2023	04/08/2023
							F1905	Device Revision or Replacement	A0908	Incorrect, Inadequate or Imprecise Result or Readings	04/08/2023	04/08/2023
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A0908	Incorrect, Inadequate or Imprecise Result or Readings	04/08/2023	04/08/2023
							F08	Hospitalization or Prolonged Hospitalization	A0908	Incorrect, Inadequate or Imprecise Result or Readings	04/08/2023	04/08/2023
							F1905	Device Revision or Replacement	A0908	Incorrect, Inadequate or Imprecise Result or Readings	04/08/2023	04/08/2023
					CT2727	Active implant devices/ancillaries	E2403	No Clinical Signs, Symptoms or Conditions	A0908	Incorrect, Inadequate or Imprecise Result or Readings	04/08/2023	04/08/2023
							F08	Hospitalization or Prolonged Hospitalization	A0908	Incorrect, Inadequate or Imprecise Result or Readings	04/08/2023	04/08/2023
							F1905	Device Revision or Replacement	A0908	Incorrect, Inadequate or Imprecise Result or Readings	04/08/2023	04/08/2023
					CT2750	Implantable cardiac electrical stimulators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A0908	Incorrect, Inadequate or Imprecise Result or Readings	04/08/2023	04/08/2023
							F08	Hospitalization or Prolonged Hospitalization	A0908	Incorrect, Inadequate or Imprecise Result or Readings	04/08/2023	04/08/2023
							F1905	Device Revision or Replacement	A0908	Incorrect, Inadequate or Imprecise Result or Readings	04/08/2023	04/08/2023
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A0908	Incorrect, Inadequate or Imprecise Result or Readings	04/08/2023	04/08/2023
							F08	Hospitalization or Prolonged Hospitalization	A0908	Incorrect, Inadequate or Imprecise Result or Readings	04/08/2023	04/08/2023
							F1905	Device Revision or Replacement	A0908	Incorrect, Inadequate or Imprecise Result or Readings	04/08/2023	04/08/2023

					CT2985	Active	E2403	No Clinical Signs, Symptoms or Conditions	A0908	Incorrect, Inadequate or Imprecise Result or Readings	04/08/2023	04/08/2023
							F08	Hospitalization or Prolonged Hospitalization	A0908	Incorrect, Inadequate or Imprecise Result or Readings	04/08/2023	04/08/2023
							F1905	Device Revision or Replacement	A0908	Incorrect, Inadequate or Imprecise Result or Readings	04/08/2023	04/08/2023
					CT3001	Implantable metal	E2403	No Clinical Signs, Symptoms or Conditions	A0908	Incorrect, Inadequate or Imprecise Result or Readings	04/08/2023	04/08/2023
							F08	Hospitalization or Prolonged Hospitalization	A0908	Incorrect, Inadequate or Imprecise Result or Readings	04/08/2023	04/08/2023
							F1905	Device Revision or Replacement	A0908	Incorrect, Inadequate or Imprecise Result or Readings	04/08/2023	04/08/2023
					CT3002	Implantable synthetic polymer	E2403	No Clinical Signs, Symptoms or Conditions	A0908	Incorrect, Inadequate or Imprecise Result or Readings	04/08/2023	04/08/2023
							F08	Hospitalization or Prolonged Hospitalization	A0908	Incorrect, Inadequate or Imprecise Result or Readings	04/08/2023	04/08/2023
							F1905	Device Revision or Replacement	A0908	Incorrect, Inadequate or Imprecise Result or Readings	04/08/2023	04/08/2023
					CT3006	Implantable inorganic material	E2403	No Clinical Signs, Symptoms or Conditions	A0908	Incorrect, Inadequate or Imprecise Result or Readings	04/08/2023	04/08/2023
							F08	Hospitalization or Prolonged Hospitalization	A0908	Incorrect, Inadequate or Imprecise Result or Readings	04/08/2023	04/08/2023
							F1905	Device Revision or Replacement	A0908	Incorrect, Inadequate or Imprecise Result or Readings	04/08/2023	04/08/2023
					CT314	Electrically powered	E2403	No Clinical Signs, Symptoms or Conditions	A0908	Incorrect, Inadequate or Imprecise Result or Readings	04/08/2023	04/08/2023
							F08	Hospitalization or Prolonged Hospitalization	A0908	Incorrect, Inadequate or Imprecise Result or Readings	04/08/2023	04/08/2023
							F1905	Device Revision or Replacement	A0908	Incorrect, Inadequate or Imprecise Result or Readings	04/08/2023	04/08/2023
					CT321	Long-term surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A0908	Incorrect, Inadequate or Imprecise Result or Readings	04/08/2023	04/08/2023
							F08	Hospitalization or Prolonged Hospitalization	A0908	Incorrect, Inadequate or Imprecise Result or Readings	04/08/2023	04/08/2023
							F1905	Device Revision or Replacement	A0908	Incorrect, Inadequate or Imprecise Result or Readings	04/08/2023	04/08/2023
					CT449	Generators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A0908	Incorrect, Inadequate or Imprecise Result or Readings	04/08/2023	04/08/2023
							F08	Hospitalization or Prolonged Hospitalization	A0908	Incorrect, Inadequate or Imprecise Result or Readings	04/08/2023	04/08/2023
							F1905	Device Revision or Replacement	A0908	Incorrect, Inadequate or Imprecise Result or Readings	04/08/2023	04/08/2023
					CT981	Single use	E2403	No Clinical Signs, Symptoms or Conditions	A0908	Incorrect, Inadequate or Imprecise Result or Readings	04/08/2023	04/08/2023
							F08	Hospitalization or Prolonged Hospitalization	A0908	Incorrect, Inadequate or Imprecise Result or Readings	04/08/2023	04/08/2023
							F1905	Device Revision or Replacement	A0908	Incorrect, Inadequate or Imprecise Result or Readings	04/08/2023	04/08/2023
					CT983	Surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A0908	Incorrect, Inadequate or Imprecise Result or Readings	04/08/2023	04/08/2023
							F08	Hospitalization or Prolonged Hospitalization	A0908	Incorrect, Inadequate or Imprecise Result or Readings	04/08/2023	04/08/2023
							F1905	Device Revision or Replacement	A0908	Incorrect, Inadequate or Imprecise Result or Readings	04/08/2023	04/08/2023
13855	2023	All other reportable incidents	35852	Single-chamber implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
									A0908	Incorrect, Inadequate or Imprecise Result or Readings		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
									A0908	Incorrect, Inadequate or Imprecise Result or Readings		
13856	2023	All other reportable incidents	47265	Dual-chamber implantable pacemaker, rate-responsive	CT141	Electrodes/Leads and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	27/09/2023	27/09/2023
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	27/09/2023	27/09/2023
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	27/09/2023	27/09/2023
					CT145	Cardiovascular devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	27/09/2023	27/09/2023
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	27/09/2023	27/09/2023
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	27/09/2023	27/09/2023
					CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	27/09/2023	27/09/2023
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	27/09/2023	27/09/2023
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	27/09/2023	27/09/2023
					CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	27/09/2023	27/09/2023
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	27/09/2023	27/09/2023
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	27/09/2023	27/09/2023
					CT233	Surgical	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	27/09/2023	27/09/2023
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	27/09/2023	27/09/2023
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	27/09/2023	27/09/2023
					CT2406	Implantable	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	27/09/2023	27/09/2023
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	27/09/2023	27/09/2023
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	27/09/2023	27/09/2023
					CT2407	Active implantable	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	27/09/2023	27/09/2023
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	27/09/2023	27/09/2023
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	27/09/2023	27/09/2023
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	27/09/2023	27/09/2023
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	27/09/2023	27/09/2023
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	27/09/2023	27/09/2023
					CT2727	Active implant devices/ancillaries	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	27/09/2023	27/09/2023
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	27/09/2023	27/09/2023
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	27/09/2023	27/09/2023
					CT2750	Implantable cardiac electrical stimulators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	27/09/2023	27/09/2023
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	27/09/2023	27/09/2023
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	27/09/2023	27/09/2023
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	27/09/2023	27/09/2023
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	27/09/2023	27/09/2023
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	27/09/2023	27/09/2023
					CT2985	Active	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	27/09/2023	27/09/2023
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	27/09/2023	27/09/2023

						F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	27/09/2023	27/09/2023	
					CT3001	Implantable metal	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	27/09/2023	27/09/2023
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	27/09/2023	27/09/2023
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	27/09/2023	27/09/2023
					CT3002	Implantable synthetic polymer	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	27/09/2023	27/09/2023
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	27/09/2023	27/09/2023
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	27/09/2023	27/09/2023
					CT3006	Implantable inorganic material	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	27/09/2023	27/09/2023
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	27/09/2023	27/09/2023
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	27/09/2023	27/09/2023
					CT314	Electrically powered	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	27/09/2023	27/09/2023
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	27/09/2023	27/09/2023
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	27/09/2023	27/09/2023
					CT321	Long-term surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	27/09/2023	27/09/2023
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	27/09/2023	27/09/2023
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	27/09/2023	27/09/2023
					CT449	Generators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	27/09/2023	27/09/2023
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	27/09/2023	27/09/2023
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	27/09/2023	27/09/2023
					CT981	Single use	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	27/09/2023	27/09/2023
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	27/09/2023	27/09/2023
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	27/09/2023	27/09/2023
					CT983	Surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	27/09/2023	27/09/2023
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	27/09/2023	27/09/2023
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	27/09/2023	27/09/2023
13857	2023	Unanticipated serious deterioration in state of health	47265	Dual-chamber implantable pacemaker, rate-responsive	CT1685	Cardiac pacemakers	E011903	Syncope/Fainting	A070101	Failure to Capture	26/06/2023	
							F1903	Device Explantation	A070101	Failure to Capture	26/06/2023	
					CT257	Implantable cardiac pacemakers	E011903	Syncope/Fainting	A070101	Failure to Capture	26/06/2023	
							F1903	Device Explantation	A070101	Failure to Capture	26/06/2023	
					CT2866	Implantable cardiac electrical stimulators	E011903	Syncope/Fainting	A070101	Failure to Capture	26/06/2023	
							F1903	Device Explantation	A070101	Failure to Capture	26/06/2023	
13858	2023	All other reportable incidents	47265	Dual-chamber implantable pacemaker, rate-responsive	CT141	Electrodes/Leads and associated devices	E2311	Discomfort	A072201	High Impedance		
							F26	No Health Consequences or Impact	A072201	High Impedance		
					CT145	Cardiovascular devices	E2311	Discomfort	A072201	High Impedance		
							F26	No Health Consequences or Impact	A072201	High Impedance		
					CT1684	Cardiac pacemakers and associated devices	E2311	Discomfort	A072201	High Impedance		
							F26	No Health Consequences or Impact	A072201	High Impedance		
					CT1685	Cardiac pacemakers	E2311	Discomfort	A072201	High Impedance		
							F26	No Health Consequences or Impact	A072201	High Impedance		
					CT233	Surgical	E2311	Discomfort	A072201	High Impedance		
							F26	No Health Consequences or Impact	A072201	High Impedance		
					CT2406	Implantable	E2311	Discomfort	A072201	High Impedance		
							F26	No Health Consequences or Impact	A072201	High Impedance		
					CT2407	Active implantable	E2311	Discomfort	A072201	High Impedance		
							F26	No Health Consequences or Impact	A072201	High Impedance		
					CT257	Implantable cardiac pacemakers	E2311	Discomfort	A072201	High Impedance		
							F26	No Health Consequences or Impact	A072201	High Impedance		
					CT2727	Active implant devices/ancillaries	E2311	Discomfort	A072201	High Impedance		
							F26	No Health Consequences or Impact	A072201	High Impedance		
					CT2750	Implantable cardiac electrical stimulators and associated devices	E2311	Discomfort	A072201	High Impedance		
							F26	No Health Consequences or Impact	A072201	High Impedance		
					CT2866	Implantable cardiac electrical stimulators	E2311	Discomfort	A072201	High Impedance		
							F26	No Health Consequences or Impact	A072201	High Impedance		
					CT2985	Active	E2311	Discomfort	A072201	High Impedance		
							F26	No Health Consequences or Impact	A072201	High Impedance		
					CT3001	Implantable metal	E2311	Discomfort	A072201	High Impedance		
							F26	No Health Consequences or Impact	A072201	High Impedance		
					CT3002	Implantable synthetic polymer	E2311	Discomfort	A072201	High Impedance		
							F26	No Health Consequences or Impact	A072201	High Impedance		
					CT3006	Implantable inorganic material	E2311	Discomfort	A072201	High Impedance		
							F26	No Health Consequences or Impact	A072201	High Impedance		
					CT314	Electrically powered	E2311	Discomfort	A072201	High Impedance		
							F26	No Health Consequences or Impact	A072201	High Impedance		
					CT321	Long-term surgical invasive	E2311	Discomfort	A072201	High Impedance		
							F26	No Health Consequences or Impact	A072201	High Impedance		
					CT449	Generators and associated devices	E2311	Discomfort	A072201	High Impedance		
							F26	No Health Consequences or Impact	A072201	High Impedance		
					CT981	Single use	E2311	Discomfort	A072201	High Impedance		
							F26	No Health Consequences or Impact	A072201	High Impedance		
					CT983	Surgical invasive	E2311	Discomfort	A072201	High Impedance		
							F26	No Health Consequences or Impact	A072201	High Impedance		
13859	2023	All other reportable incidents	17882	Physiologic-monitoring defibrillation system	CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A0708	Power Problem		
							F27	Problem Identified During Non-Clinical Procedure	A0708	Power Problem		
13860	2023	All other reportable incidents	47265	Dual-chamber implantable pacemaker, rate-responsive	CT141	Electrodes/Leads and associated devices	E0112	Dizziness	A070101	Failure to Capture		
							E011903	Syncope/Fainting	A070101	Failure to Capture		
							E060101	Asystole	A070101	Failure to Capture		
							E2007	Fall	A070101	Failure to Capture		
							F08	Hospitalization or Prolonged Hospitalization	A070101	Failure to Capture		
							F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A070101	Failure to Capture		
					CT145	Cardiovascular devices	E0112	Dizziness	A070101	Failure to Capture		
							E011903	Syncope/Fainting	A070101	Failure to Capture		
							E060101	Asystole	A070101	Failure to Capture		
							E2007	Fall	A070101	Failure to Capture		
							F08	Hospitalization or Prolonged Hospitalization	A070101	Failure to Capture		
							F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A070101	Failure to Capture		
					CT1684	Cardiac pacemakers and associated devices	E0112	Dizziness	A070101	Failure to Capture		
							E011903	Syncope/Fainting	A070101	Failure to Capture		
							E060101	Asystole	A070101	Failure to Capture		

							E2007	Fall	A070101	Failure to Capture		
							F08	Hospitalization or Prolonged Hospitalization	A070101	Failure to Capture		
							F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A070101	Failure to Capture		
						CT1685	Cardiac pacemakers	E0112	Dizziness	A070101	Failure to Capture		
								E011903	Syncope/Fainting	A070101	Failure to Capture		
								E060101	Asystole	A070101	Failure to Capture		
								E2007	Fall	A070101	Failure to Capture		
								F08	Hospitalization or Prolonged Hospitalization	A070101	Failure to Capture		
								F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A070101	Failure to Capture		
						CT233	Surgical	E0112	Dizziness	A070101	Failure to Capture		
								E011903	Syncope/Fainting	A070101	Failure to Capture		
								E060101	Asystole	A070101	Failure to Capture		
								E2007	Fall	A070101	Failure to Capture		
								F08	Hospitalization or Prolonged Hospitalization	A070101	Failure to Capture		
								F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A070101	Failure to Capture		
						CT2406	Implantable	E0112	Dizziness	A070101	Failure to Capture		
								E011903	Syncope/Fainting	A070101	Failure to Capture		
								E060101	Asystole	A070101	Failure to Capture		
								E2007	Fall	A070101	Failure to Capture		
								F08	Hospitalization or Prolonged Hospitalization	A070101	Failure to Capture		
								F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A070101	Failure to Capture		
						CT2407	Active implantable	E0112	Dizziness	A070101	Failure to Capture		
								E011903	Syncope/Fainting	A070101	Failure to Capture		
								E060101	Asystole	A070101	Failure to Capture		
								E2007	Fall	A070101	Failure to Capture		
								F08	Hospitalization or Prolonged Hospitalization	A070101	Failure to Capture		
								F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A070101	Failure to Capture		
						CT257	Implantable cardiac pacemakers	E0112	Dizziness	A070101	Failure to Capture		
								E011903	Syncope/Fainting	A070101	Failure to Capture		
								E060101	Asystole	A070101	Failure to Capture		
								E2007	Fall	A070101	Failure to Capture		
								F08	Hospitalization or Prolonged Hospitalization	A070101	Failure to Capture		
								F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A070101	Failure to Capture		
						CT2727	Active implant devices/ancillaries	E0112	Dizziness	A070101	Failure to Capture		
								E011903	Syncope/Fainting	A070101	Failure to Capture		
								E060101	Asystole	A070101	Failure to Capture		
								E2007	Fall	A070101	Failure to Capture		
								F08	Hospitalization or Prolonged Hospitalization	A070101	Failure to Capture		
								F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A070101	Failure to Capture		
						CT2750	Implantable cardiac electrical stimulators and associated devices	E0112	Dizziness	A070101	Failure to Capture		
								E011903	Syncope/Fainting	A070101	Failure to Capture		
								E060101	Asystole	A070101	Failure to Capture		
								E2007	Fall	A070101	Failure to Capture		
								F08	Hospitalization or Prolonged Hospitalization	A070101	Failure to Capture		
								F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A070101	Failure to Capture		
						CT2866	Implantable cardiac electrical stimulators	E0112	Dizziness	A070101	Failure to Capture		
								E011903	Syncope/Fainting	A070101	Failure to Capture		
								E060101	Asystole	A070101	Failure to Capture		
								E2007	Fall	A070101	Failure to Capture		
								F08	Hospitalization or Prolonged Hospitalization	A070101	Failure to Capture		
								F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A070101	Failure to Capture		
						CT2985	Active	E0112	Dizziness	A070101	Failure to Capture		
								E011903	Syncope/Fainting	A070101	Failure to Capture		
								E060101	Asystole	A070101	Failure to Capture		
								E2007	Fall	A070101	Failure to Capture		
								F08	Hospitalization or Prolonged Hospitalization	A070101	Failure to Capture		
								F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A070101	Failure to Capture		
						CT3001	Implantable metal	E0112	Dizziness	A070101	Failure to Capture		
								E011903	Syncope/Fainting	A070101	Failure to Capture		
								E060101	Asystole	A070101	Failure to Capture		
								E2007	Fall	A070101	Failure to Capture		
								F08	Hospitalization or Prolonged Hospitalization	A070101	Failure to Capture		
								F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A070101	Failure to Capture		
						CT3002	Implantable synthetic polymer	E0112	Dizziness	A070101	Failure to Capture		
								E011903	Syncope/Fainting	A070101	Failure to Capture		
								E060101	Asystole	A070101	Failure to Capture		
								E2007	Fall	A070101	Failure to Capture		
								F08	Hospitalization or Prolonged Hospitalization	A070101	Failure to Capture		
								F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A070101	Failure to Capture		
						CT3006	Implantable inorganic material	E0112	Dizziness	A070101	Failure to Capture		
								E011903	Syncope/Fainting	A070101	Failure to Capture		
								E060101	Asystole	A070101	Failure to Capture		
								E2007	Fall	A070101	Failure to Capture		
								F08	Hospitalization or Prolonged Hospitalization	A070101	Failure to Capture		
								F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A070101	Failure to Capture		
						CT314	Electrically powered	E0112	Dizziness	A070101	Failure to Capture		
								E011903	Syncope/Fainting	A070101	Failure to Capture		
								E060101	Asystole	A070101	Failure to Capture		
								E2007	Fall	A070101	Failure to Capture		
								F08	Hospitalization or Prolonged Hospitalization	A070101	Failure to Capture		
								F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A070101	Failure to Capture		
						CT321	Long-term surgical invasive	E0112	Dizziness	A070101	Failure to Capture		
								E011903	Syncope/Fainting	A070101	Failure to Capture		
								E060101	Asystole	A070101	Failure to Capture		
								E2007	Fall	A070101	Failure to Capture		
								F08	Hospitalization or Prolonged Hospitalization	A070101	Failure to Capture		
								F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A070101	Failure to Capture		
						CT449	Generators and associated devices	E0112	Dizziness	A070101	Failure to Capture		

							E011903	Syncope/Fainting	A070101	Failure to Capture		
							E060101	Asystole	A070101	Failure to Capture		
							E2007	Fall	A070101	Failure to Capture		
							F08	Hospitalization or Prolonged Hospitalization	A070101	Failure to Capture		
							F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A070101	Failure to Capture		
					CT981	Single use	E0112	Dizziness	A070101	Failure to Capture		
							E011903	Syncope/Fainting	A070101	Failure to Capture		
							E060101	Asystole	A070101	Failure to Capture		
							E2007	Fall	A070101	Failure to Capture		
							F08	Hospitalization or Prolonged Hospitalization	A070101	Failure to Capture		
							F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A070101	Failure to Capture		
					CT983	Surgical invasive	E0112	Dizziness	A070101	Failure to Capture		
							E011903	Syncope/Fainting	A070101	Failure to Capture		
							E060101	Asystole	A070101	Failure to Capture		
							E2007	Fall	A070101	Failure to Capture		
							F08	Hospitalization or Prolonged Hospitalization	A070101	Failure to Capture		
							F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A070101	Failure to Capture		
13861	2023	All other reportable incidents	36241	Pacemaker/defibrillator lead	CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070910	Undersensing		
									A1301	Failure to Read Input Signal		
							F26	No Health Consequences or Impact	A070910	Undersensing		
									A1301	Failure to Read Input Signal		
13862	2023	All other reportable incidents	47265	Dual-chamber implantable pacemaker, rate-responsive	CT141	Electrodes/Leads and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
					CT145	Cardiovascular devices	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
					CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
					CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
					CT233	Surgical	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
					CT2406	Implantable	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
					CT2407	Active implantable	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
					CT2727	Active implant devices/ancillaries	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
					CT2750	Implantable cardiac electrical stimulators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
					CT2985	Active	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
					CT3001	Implantable metal	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
					CT3002	Implantable synthetic polymer	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
					CT3006	Implantable inorganic material	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
					CT314	Electrically powered	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
					CT321	Long-term surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A0705	Battery Problem		

									A070504	Premature Discharge of Battery		
					CT449	Generators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
					CT981	Single use	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
					CT983	Surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
13863	2023	All other reportable incidents	47265	Dual-chamber implantable pacemaker, rate-responsive	CT141	Electrodes/Leads and associated devices	E1020	Nausea	A04	Material Integrity Problem	12/07/2023	12/07/2023
									A070101	Failure to Capture	12/07/2023	12/07/2023
							E2109	Implant Pain	A04	Material Integrity Problem	12/07/2023	12/07/2023
									A070101	Failure to Capture	12/07/2023	12/07/2023
							F08	Hospitalization or Prolonged Hospitalization	A04	Material Integrity Problem	12/07/2023	12/07/2023
									A070101	Failure to Capture	12/07/2023	12/07/2023
							F11	Non-Serious Injury/ Illness/ Impairment	A04	Material Integrity Problem	12/07/2023	12/07/2023
									A070101	Failure to Capture	12/07/2023	12/07/2023
							F1903	Device Explantation	A04	Material Integrity Problem	12/07/2023	12/07/2023
									A070101	Failure to Capture	12/07/2023	12/07/2023
					CT145	Cardiovascular devices	E1020	Nausea	A04	Material Integrity Problem	12/07/2023	12/07/2023
									A070101	Failure to Capture	12/07/2023	12/07/2023
							E2109	Implant Pain	A04	Material Integrity Problem	12/07/2023	12/07/2023
									A070101	Failure to Capture	12/07/2023	12/07/2023
							F08	Hospitalization or Prolonged Hospitalization	A04	Material Integrity Problem	12/07/2023	12/07/2023
									A070101	Failure to Capture	12/07/2023	12/07/2023
							F11	Non-Serious Injury/ Illness/ Impairment	A04	Material Integrity Problem	12/07/2023	12/07/2023
									A070101	Failure to Capture	12/07/2023	12/07/2023
							F1903	Device Explantation	A04	Material Integrity Problem	12/07/2023	12/07/2023
									A070101	Failure to Capture	12/07/2023	12/07/2023
					CT1684	Cardiac pacemakers and associated devices	E1020	Nausea	A04	Material Integrity Problem	12/07/2023	12/07/2023
									A070101	Failure to Capture	12/07/2023	12/07/2023
							E2109	Implant Pain	A04	Material Integrity Problem	12/07/2023	12/07/2023
									A070101	Failure to Capture	12/07/2023	12/07/2023
							F08	Hospitalization or Prolonged Hospitalization	A04	Material Integrity Problem	12/07/2023	12/07/2023
									A070101	Failure to Capture	12/07/2023	12/07/2023
							F11	Non-Serious Injury/ Illness/ Impairment	A04	Material Integrity Problem	12/07/2023	12/07/2023
									A070101	Failure to Capture	12/07/2023	12/07/2023
							F1903	Device Explantation	A04	Material Integrity Problem	12/07/2023	12/07/2023
									A070101	Failure to Capture	12/07/2023	12/07/2023
					CT1685	Cardiac pacemakers	E1020	Nausea	A04	Material Integrity Problem	12/07/2023	12/07/2023
									A070101	Failure to Capture	12/07/2023	12/07/2023
							E2109	Implant Pain	A04	Material Integrity Problem	12/07/2023	12/07/2023
									A070101	Failure to Capture	12/07/2023	12/07/2023
							F08	Hospitalization or Prolonged Hospitalization	A04	Material Integrity Problem	12/07/2023	12/07/2023
									A070101	Failure to Capture	12/07/2023	12/07/2023
							F11	Non-Serious Injury/ Illness/ Impairment	A04	Material Integrity Problem	12/07/2023	12/07/2023
									A070101	Failure to Capture	12/07/2023	12/07/2023
							F1903	Device Explantation	A04	Material Integrity Problem	12/07/2023	12/07/2023
									A070101	Failure to Capture	12/07/2023	12/07/2023
					CT233	Surgical	E1020	Nausea	A04	Material Integrity Problem	12/07/2023	12/07/2023
									A070101	Failure to Capture	12/07/2023	12/07/2023
							E2109	Implant Pain	A04	Material Integrity Problem	12/07/2023	12/07/2023
									A070101	Failure to Capture	12/07/2023	12/07/2023
							F08	Hospitalization or Prolonged Hospitalization	A04	Material Integrity Problem	12/07/2023	12/07/2023
									A070101	Failure to Capture	12/07/2023	12/07/2023
							F11	Non-Serious Injury/ Illness/ Impairment	A04	Material Integrity Problem	12/07/2023	12/07/2023
									A070101	Failure to Capture	12/07/2023	12/07/2023
							F1903	Device Explantation	A04	Material Integrity Problem	12/07/2023	12/07/2023
									A070101	Failure to Capture	12/07/2023	12/07/2023
					CT2406	Implantable	E1020	Nausea	A04	Material Integrity Problem	12/07/2023	12/07/2023
									A070101	Failure to Capture	12/07/2023	12/07/2023
							E2109	Implant Pain	A04	Material Integrity Problem	12/07/2023	12/07/2023
									A070101	Failure to Capture	12/07/2023	12/07/2023
							F08	Hospitalization or Prolonged Hospitalization	A04	Material Integrity Problem	12/07/2023	12/07/2023
									A070101	Failure to Capture	12/07/2023	12/07/2023
							F11	Non-Serious Injury/ Illness/ Impairment	A04	Material Integrity Problem	12/07/2023	12/07/2023
									A070101	Failure to Capture	12/07/2023	12/07/2023
							F1903	Device Explantation	A04	Material Integrity Problem	12/07/2023	12/07/2023
									A070101	Failure to Capture	12/07/2023	12/07/2023
					CT2407	Active implantable	E1020	Nausea	A04	Material Integrity Problem	12/07/2023	12/07/2023
									A070101	Failure to Capture	12/07/2023	12/07/2023
							E2109	Implant Pain	A04	Material Integrity Problem	12/07/2023	12/07/2023
									A070101	Failure to Capture	12/07/2023	12/07/2023
							F08	Hospitalization or Prolonged Hospitalization	A04	Material Integrity Problem	12/07/2023	12/07/2023
									A070101	Failure to Capture	12/07/2023	12/07/2023
							F11	Non-Serious Injury/ Illness/ Impairment	A04	Material Integrity Problem	12/07/2023	12/07/2023
									A070101	Failure to Capture	12/07/2023	12/07/2023
							F1903	Device Explantation	A04	Material Integrity Problem	12/07/2023	12/07/2023
									A070101	Failure to Capture	12/07/2023	12/07/2023
					CT257	Implantable cardiac pacemakers	E1020	Nausea	A04	Material Integrity Problem	12/07/2023	12/07/2023
									A070101	Failure to Capture	12/07/2023	12/07/2023
							E2109	Implant Pain	A04	Material Integrity Problem	12/07/2023	12/07/2023
									A070101	Failure to Capture	12/07/2023	12/07/2023
							F08	Hospitalization or Prolonged Hospitalization	A04	Material Integrity Problem	12/07/2023	12/07/2023

								A070101	Failure to Capture	12/07/2023	12/07/2023
							F11	Non-Serious Injury/ Illness/ Impairment	A04	Material Integrity Problem	12/07/2023	12/07/2023
									A070101	Failure to Capture	12/07/2023	12/07/2023
							F1903	Device Explantation	A04	Material Integrity Problem	12/07/2023	12/07/2023
									A070101	Failure to Capture	12/07/2023	12/07/2023
					CT2727	Active implant devices/ancillaries	E1020	Nausea	A04	Material Integrity Problem	12/07/2023	12/07/2023
									A070101	Failure to Capture	12/07/2023	12/07/2023
							E2109	Implant Pain	A04	Material Integrity Problem	12/07/2023	12/07/2023
									A070101	Failure to Capture	12/07/2023	12/07/2023
							F08	Hospitalization or Prolonged Hospitalization	A04	Material Integrity Problem	12/07/2023	12/07/2023
									A070101	Failure to Capture	12/07/2023	12/07/2023
							F11	Non-Serious Injury/ Illness/ Impairment	A04	Material Integrity Problem	12/07/2023	12/07/2023
									A070101	Failure to Capture	12/07/2023	12/07/2023
							F1903	Device Explantation	A04	Material Integrity Problem	12/07/2023	12/07/2023
									A070101	Failure to Capture	12/07/2023	12/07/2023
					CT2750	Implantable cardiac electrical stimulators and associated devices	E1020	Nausea	A04	Material Integrity Problem	12/07/2023	12/07/2023
									A070101	Failure to Capture	12/07/2023	12/07/2023
							E2109	Implant Pain	A04	Material Integrity Problem	12/07/2023	12/07/2023
									A070101	Failure to Capture	12/07/2023	12/07/2023
							F08	Hospitalization or Prolonged Hospitalization	A04	Material Integrity Problem	12/07/2023	12/07/2023
									A070101	Failure to Capture	12/07/2023	12/07/2023
							F11	Non-Serious Injury/ Illness/ Impairment	A04	Material Integrity Problem	12/07/2023	12/07/2023
									A070101	Failure to Capture	12/07/2023	12/07/2023
							F1903	Device Explantation	A04	Material Integrity Problem	12/07/2023	12/07/2023
									A070101	Failure to Capture	12/07/2023	12/07/2023
					CT2866	Implantable cardiac electrical stimulators	E1020	Nausea	A04	Material Integrity Problem	12/07/2023	12/07/2023
									A070101	Failure to Capture	12/07/2023	12/07/2023
							E2109	Implant Pain	A04	Material Integrity Problem	12/07/2023	12/07/2023
									A070101	Failure to Capture	12/07/2023	12/07/2023
							F08	Hospitalization or Prolonged Hospitalization	A04	Material Integrity Problem	12/07/2023	12/07/2023
									A070101	Failure to Capture	12/07/2023	12/07/2023
							F11	Non-Serious Injury/ Illness/ Impairment	A04	Material Integrity Problem	12/07/2023	12/07/2023
									A070101	Failure to Capture	12/07/2023	12/07/2023
							F1903	Device Explantation	A04	Material Integrity Problem	12/07/2023	12/07/2023
									A070101	Failure to Capture	12/07/2023	12/07/2023
					CT2985	Active	E1020	Nausea	A04	Material Integrity Problem	12/07/2023	12/07/2023
									A070101	Failure to Capture	12/07/2023	12/07/2023
							E2109	Implant Pain	A04	Material Integrity Problem	12/07/2023	12/07/2023
									A070101	Failure to Capture	12/07/2023	12/07/2023
							F08	Hospitalization or Prolonged Hospitalization	A04	Material Integrity Problem	12/07/2023	12/07/2023
									A070101	Failure to Capture	12/07/2023	12/07/2023
							F11	Non-Serious Injury/ Illness/ Impairment	A04	Material Integrity Problem	12/07/2023	12/07/2023
									A070101	Failure to Capture	12/07/2023	12/07/2023
							F1903	Device Explantation	A04	Material Integrity Problem	12/07/2023	12/07/2023
									A070101	Failure to Capture	12/07/2023	12/07/2023
					CT3001	Implantable metal	E1020	Nausea	A04	Material Integrity Problem	12/07/2023	12/07/2023
									A070101	Failure to Capture	12/07/2023	12/07/2023
							E2109	Implant Pain	A04	Material Integrity Problem	12/07/2023	12/07/2023
									A070101	Failure to Capture	12/07/2023	12/07/2023
							F08	Hospitalization or Prolonged Hospitalization	A04	Material Integrity Problem	12/07/2023	12/07/2023
									A070101	Failure to Capture	12/07/2023	12/07/2023
							F11	Non-Serious Injury/ Illness/ Impairment	A04	Material Integrity Problem	12/07/2023	12/07/2023
									A070101	Failure to Capture	12/07/2023	12/07/2023
							F1903	Device Explantation	A04	Material Integrity Problem	12/07/2023	12/07/2023
									A070101	Failure to Capture	12/07/2023	12/07/2023
					CT3002	Implantable synthetic polymer	E1020	Nausea	A04	Material Integrity Problem	12/07/2023	12/07/2023
									A070101	Failure to Capture	12/07/2023	12/07/2023
							E2109	Implant Pain	A04	Material Integrity Problem	12/07/2023	12/07/2023
									A070101	Failure to Capture	12/07/2023	12/07/2023
							F08	Hospitalization or Prolonged Hospitalization	A04	Material Integrity Problem	12/07/2023	12/07/2023
									A070101	Failure to Capture	12/07/2023	12/07/2023
							F11	Non-Serious Injury/ Illness/ Impairment	A04	Material Integrity Problem	12/07/2023	12/07/2023
									A070101	Failure to Capture	12/07/2023	12/07/2023
							F1903	Device Explantation	A04	Material Integrity Problem	12/07/2023	12/07/2023
									A070101	Failure to Capture	12/07/2023	12/07/2023
					CT3006	Implantable inorganic material	E1020	Nausea	A04	Material Integrity Problem	12/07/2023	12/07/2023
									A070101	Failure to Capture	12/07/2023	12/07/2023
							E2109	Implant Pain	A04	Material Integrity Problem	12/07/2023	12/07/2023
									A070101	Failure to Capture	12/07/2023	12/07/2023
							F08	Hospitalization or Prolonged Hospitalization	A04	Material Integrity Problem	12/07/2023	12/07/2023
									A070101	Failure to Capture	12/07/2023	12/07/2023
							F11	Non-Serious Injury/ Illness/ Impairment	A04	Material Integrity Problem	12/07/2023	12/07/2023
									A070101	Failure to Capture	12/07/2023	12/07/2023
							F1903	Device Explantation	A04	Material Integrity Problem	12/07/2023	12/07/2023
									A070101	Failure to Capture	12/07/2023	12/07/2023
					CT314	Electrically powered	E1020	Nausea	A04	Material Integrity Problem	12/07/2023	12/07/2023
									A070101	Failure to Capture	12/07/2023	12/07/2023
							E2109	Implant Pain	A04	Material Integrity Problem	12/07/2023	12/07/2023
									A070101	Failure to Capture	12/07/2023	12/07/2023
							F08	Hospitalization or Prolonged Hospitalization	A04	Material Integrity Problem	12/07/2023	12/07/2023
									A070101	Failure to Capture	12/07/2023	12/07/2023
							F11	Non-Serious Injury/ Illness/ Impairment	A04	Material Integrity Problem	12/07/2023	12/07/2023
									A070101	Failure to Capture	12/07/2023	12/07/2023
							F1903	Device Explantation	A04	Material Integrity Problem	12/07/2023	12/07/2023
									A070101	Failure to Capture	12/07/2023	12/07/2023
					CT321	Long-term surgical invasive	E1020	Nausea	A04	Material Integrity Problem	12/07/2023	12/07/2023
									A070101	Failure to Capture	12/07/2023	12/07/2023
							E2109	Implant Pain	A04	Material Integrity Problem	12/07/2023	12/07/2023

									A070101	Failure to Capture	12/07/2023	12/07/2023
							F08	Hospitalization or Prolonged Hospitalization	A04	Material Integrity Problem	12/07/2023	12/07/2023
									A070101	Failure to Capture	12/07/2023	12/07/2023
							F11	Non-Serious Injury/ Illness/ Impairment	A04	Material Integrity Problem	12/07/2023	12/07/2023
									A070101	Failure to Capture	12/07/2023	12/07/2023
							F1903	Device Explantation	A04	Material Integrity Problem	12/07/2023	12/07/2023
									A070101	Failure to Capture	12/07/2023	12/07/2023
					CT449	Generators and associated devices	E1020	Nausea	A04	Material Integrity Problem	12/07/2023	12/07/2023
									A070101	Failure to Capture	12/07/2023	12/07/2023
							E2109	Implant Pain	A04	Material Integrity Problem	12/07/2023	12/07/2023
									A070101	Failure to Capture	12/07/2023	12/07/2023
							F08	Hospitalization or Prolonged Hospitalization	A04	Material Integrity Problem	12/07/2023	12/07/2023
									A070101	Failure to Capture	12/07/2023	12/07/2023
							F11	Non-Serious Injury/ Illness/ Impairment	A04	Material Integrity Problem	12/07/2023	12/07/2023
									A070101	Failure to Capture	12/07/2023	12/07/2023
							F1903	Device Explantation	A04	Material Integrity Problem	12/07/2023	12/07/2023
									A070101	Failure to Capture	12/07/2023	12/07/2023
					CT981	Single use	E1020	Nausea	A04	Material Integrity Problem	12/07/2023	12/07/2023
									A070101	Failure to Capture	12/07/2023	12/07/2023
							E2109	Implant Pain	A04	Material Integrity Problem	12/07/2023	12/07/2023
									A070101	Failure to Capture	12/07/2023	12/07/2023
							F08	Hospitalization or Prolonged Hospitalization	A04	Material Integrity Problem	12/07/2023	12/07/2023
									A070101	Failure to Capture	12/07/2023	12/07/2023
							F11	Non-Serious Injury/ Illness/ Impairment	A04	Material Integrity Problem	12/07/2023	12/07/2023
									A070101	Failure to Capture	12/07/2023	12/07/2023
							F1903	Device Explantation	A04	Material Integrity Problem	12/07/2023	12/07/2023
									A070101	Failure to Capture	12/07/2023	12/07/2023
					CT983	Surgical invasive	E1020	Nausea	A04	Material Integrity Problem	12/07/2023	12/07/2023
									A070101	Failure to Capture	12/07/2023	12/07/2023
							E2109	Implant Pain	A04	Material Integrity Problem	12/07/2023	12/07/2023
									A070101	Failure to Capture	12/07/2023	12/07/2023
							F08	Hospitalization or Prolonged Hospitalization	A04	Material Integrity Problem	12/07/2023	12/07/2023
									A070101	Failure to Capture	12/07/2023	12/07/2023
							F11	Non-Serious Injury/ Illness/ Impairment	A04	Material Integrity Problem	12/07/2023	12/07/2023
									A070101	Failure to Capture	12/07/2023	12/07/2023
							F1903	Device Explantation	A04	Material Integrity Problem	12/07/2023	12/07/2023
									A070101	Failure to Capture	12/07/2023	12/07/2023
13864	2023	All other reportable incidents	35852	Single-chamber implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	26/10/2023	26/10/2023
									A0908	Incorrect, Inadequate or Imprecise Result or Readings	26/10/2023	26/10/2023
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	26/10/2023	26/10/2023
									A0908	Incorrect, Inadequate or Imprecise Result or Readings	26/10/2023	26/10/2023
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	26/10/2023	26/10/2023
									A0908	Incorrect, Inadequate or Imprecise Result or Readings	26/10/2023	26/10/2023
13865	2023	All other reportable incidents	17882	Physiologic-monitoring defibrillation system	CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A05	Mechanical Problem		
							F27	Problem Identified During Non-Clinical Procedure	A05	Mechanical Problem		
13866	2023	All other reportable incidents	36241	Pacemaker/defibrillator lead	CT1684	Cardiac pacemakers and associated devices	E011903	Syncope/Fainting	A090801	Signal Artifact/Noise		
							F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A090801	Signal Artifact/Noise		
							F1905	Device Revision or Replacement	A090801	Signal Artifact/Noise		
13867	2023	All other reportable incidents	36241	Pacemaker/defibrillator lead	CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070101	Failure to Capture		
									A070909	Oversensing		
									A070910	Undersensing		
									A072201	High Impedance		
									A1301	Failure to Read Input Signal		
							F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A070101	Failure to Capture		
									A070909	Oversensing		
									A070910	Undersensing		
									A072201	High Impedance		
									A1301	Failure to Read Input Signal		
13868	2023	All other reportable incidents	35852	Single-chamber implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E1906	Unspecified Infection	A24	No Apparent Device Problem		
							F08	Hospitalization or Prolonged Hospitalization	A24	No Apparent Device Problem		
							F1905	Device Revision or Replacement	A24	No Apparent Device Problem		
13869	2023	All other reportable incidents	37265	Dual-chamber implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem	15/08/2023	15/08/2023
									A070504	Premature Discharge of Battery	15/08/2023	15/08/2023
							F08	Hospitalization or Prolonged Hospitalization	A0705	Battery Problem	15/08/2023	15/08/2023
									A070504	Premature Discharge of Battery	15/08/2023	15/08/2023
							F1905	Device Revision or Replacement	A0705	Battery Problem	15/08/2023	15/08/2023
									A070504	Premature Discharge of Battery	15/08/2023	15/08/2023
13870	2023	All other reportable incidents	47265	Dual-chamber implantable pacemaker, rate-responsive	CT141	Electrodes/Leads and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
					CT145	Cardiovascular devices	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
					CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
					CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
					CT233	Surgical	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		

							F26	No Health Consequences or Impact	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
					CT2406	Implantable	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
					CT2407	Active implantable	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
					CT2727	Active implant devices/ancillaries	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
					CT2750	Implantable cardiac electrical stimulators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
					CT2985	Active	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
					CT3001	Implantable metal	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
					CT3002	Implantable synthetic polymer	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
					CT3006	Implantable inorganic material	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
					CT314	Electrically powered	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
					CT321	Long-term surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
					CT449	Generators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
					CT981	Single use	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
					CT983	Surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
13871	2023	All other reportable incidents	47265	Dual-chamber implantable pacemaker, rate-responsive	CT141	Electrodes/Leads and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem	21/02/2023	21/02/2023
									A070504	Premature Discharge of Battery	21/02/2023	21/02/2023
							F08	Hospitalization or Prolonged Hospitalization	A0705	Battery Problem	21/02/2023	21/02/2023
									A070504	Premature Discharge of Battery	21/02/2023	21/02/2023
							F1905	Device Revision or Replacement	A0705	Battery Problem	21/02/2023	21/02/2023
									A070504	Premature Discharge of Battery	21/02/2023	21/02/2023
					CT145	Cardiovascular devices	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem	21/02/2023	21/02/2023
									A070504	Premature Discharge of Battery	21/02/2023	21/02/2023
							F08	Hospitalization or Prolonged Hospitalization	A0705	Battery Problem	21/02/2023	21/02/2023
									A070504	Premature Discharge of Battery	21/02/2023	21/02/2023
							F1905	Device Revision or Replacement	A0705	Battery Problem	21/02/2023	21/02/2023
									A070504	Premature Discharge of Battery	21/02/2023	21/02/2023
					CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem	21/02/2023	21/02/2023
									A070504	Premature Discharge of Battery	21/02/2023	21/02/2023
							F08	Hospitalization or Prolonged Hospitalization	A0705	Battery Problem	21/02/2023	21/02/2023
									A070504	Premature Discharge of Battery	21/02/2023	21/02/2023
							F1905	Device Revision or Replacement	A0705	Battery Problem	21/02/2023	21/02/2023
									A070504	Premature Discharge of Battery	21/02/2023	21/02/2023
					CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem	21/02/2023	21/02/2023
									A070504	Premature Discharge of Battery	21/02/2023	21/02/2023
							F08	Hospitalization or Prolonged Hospitalization	A0705	Battery Problem	21/02/2023	21/02/2023
									A070504	Premature Discharge of Battery	21/02/2023	21/02/2023
							F1905	Device Revision or Replacement	A0705	Battery Problem	21/02/2023	21/02/2023
									A070504	Premature Discharge of Battery	21/02/2023	21/02/2023
					CT233	Surgical	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem	21/02/2023	21/02/2023
									A070504	Premature Discharge of Battery	21/02/2023	21/02/2023

		F08	Hospitalization or Prolonged Hospitalization	A0705	Battery Problem	21/02/2023	21/02/2023
				A070504	Premature Discharge of Battery	21/02/2023	21/02/2023
		F1905	Device Revision or Replacement	A0705	Battery Problem	21/02/2023	21/02/2023
				A070504	Premature Discharge of Battery	21/02/2023	21/02/2023
CT2406	Implantable	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem	21/02/2023	21/02/2023
				A070504	Premature Discharge of Battery	21/02/2023	21/02/2023
		F08	Hospitalization or Prolonged Hospitalization	A0705	Battery Problem	21/02/2023	21/02/2023
				A070504	Premature Discharge of Battery	21/02/2023	21/02/2023
		F1905	Device Revision or Replacement	A0705	Battery Problem	21/02/2023	21/02/2023
				A070504	Premature Discharge of Battery	21/02/2023	21/02/2023
CT2407	Active implantable	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem	21/02/2023	21/02/2023
				A070504	Premature Discharge of Battery	21/02/2023	21/02/2023
		F08	Hospitalization or Prolonged Hospitalization	A0705	Battery Problem	21/02/2023	21/02/2023
				A070504	Premature Discharge of Battery	21/02/2023	21/02/2023
		F1905	Device Revision or Replacement	A0705	Battery Problem	21/02/2023	21/02/2023
				A070504	Premature Discharge of Battery	21/02/2023	21/02/2023
CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem	21/02/2023	21/02/2023
				A070504	Premature Discharge of Battery	21/02/2023	21/02/2023
		F08	Hospitalization or Prolonged Hospitalization	A0705	Battery Problem	21/02/2023	21/02/2023
				A070504	Premature Discharge of Battery	21/02/2023	21/02/2023
		F1905	Device Revision or Replacement	A0705	Battery Problem	21/02/2023	21/02/2023
				A070504	Premature Discharge of Battery	21/02/2023	21/02/2023
CT2727	Active implant devices/ancillaries	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem	21/02/2023	21/02/2023
				A070504	Premature Discharge of Battery	21/02/2023	21/02/2023
		F08	Hospitalization or Prolonged Hospitalization	A0705	Battery Problem	21/02/2023	21/02/2023
				A070504	Premature Discharge of Battery	21/02/2023	21/02/2023
		F1905	Device Revision or Replacement	A0705	Battery Problem	21/02/2023	21/02/2023
				A070504	Premature Discharge of Battery	21/02/2023	21/02/2023
CT2750	Implantable cardiac electrical stimulators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem	21/02/2023	21/02/2023
				A070504	Premature Discharge of Battery	21/02/2023	21/02/2023
		F08	Hospitalization or Prolonged Hospitalization	A0705	Battery Problem	21/02/2023	21/02/2023
				A070504	Premature Discharge of Battery	21/02/2023	21/02/2023
		F1905	Device Revision or Replacement	A0705	Battery Problem	21/02/2023	21/02/2023
				A070504	Premature Discharge of Battery	21/02/2023	21/02/2023
CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem	21/02/2023	21/02/2023
				A070504	Premature Discharge of Battery	21/02/2023	21/02/2023
		F08	Hospitalization or Prolonged Hospitalization	A0705	Battery Problem	21/02/2023	21/02/2023
				A070504	Premature Discharge of Battery	21/02/2023	21/02/2023
		F1905	Device Revision or Replacement	A0705	Battery Problem	21/02/2023	21/02/2023
				A070504	Premature Discharge of Battery	21/02/2023	21/02/2023
CT2985	Active	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem	21/02/2023	21/02/2023
				A070504	Premature Discharge of Battery	21/02/2023	21/02/2023
		F08	Hospitalization or Prolonged Hospitalization	A0705	Battery Problem	21/02/2023	21/02/2023
				A070504	Premature Discharge of Battery	21/02/2023	21/02/2023
		F1905	Device Revision or Replacement	A0705	Battery Problem	21/02/2023	21/02/2023
				A070504	Premature Discharge of Battery	21/02/2023	21/02/2023
CT3001	Implantable metal	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem	21/02/2023	21/02/2023
				A070504	Premature Discharge of Battery	21/02/2023	21/02/2023
		F08	Hospitalization or Prolonged Hospitalization	A0705	Battery Problem	21/02/2023	21/02/2023
				A070504	Premature Discharge of Battery	21/02/2023	21/02/2023
		F1905	Device Revision or Replacement	A0705	Battery Problem	21/02/2023	21/02/2023
				A070504	Premature Discharge of Battery	21/02/2023	21/02/2023
CT3002	Implantable synthetic polymer	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem	21/02/2023	21/02/2023
				A070504	Premature Discharge of Battery	21/02/2023	21/02/2023
		F08	Hospitalization or Prolonged Hospitalization	A0705	Battery Problem	21/02/2023	21/02/2023
				A070504	Premature Discharge of Battery	21/02/2023	21/02/2023
		F1905	Device Revision or Replacement	A0705	Battery Problem	21/02/2023	21/02/2023
				A070504	Premature Discharge of Battery	21/02/2023	21/02/2023
CT3006	Implantable inorganic material	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem	21/02/2023	21/02/2023
				A070504	Premature Discharge of Battery	21/02/2023	21/02/2023
		F08	Hospitalization or Prolonged Hospitalization	A0705	Battery Problem	21/02/2023	21/02/2023
				A070504	Premature Discharge of Battery	21/02/2023	21/02/2023
		F1905	Device Revision or Replacement	A0705	Battery Problem	21/02/2023	21/02/2023
				A070504	Premature Discharge of Battery	21/02/2023	21/02/2023
CT314	Electrically powered	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem	21/02/2023	21/02/2023
				A070504	Premature Discharge of Battery	21/02/2023	21/02/2023
		F08	Hospitalization or Prolonged Hospitalization	A0705	Battery Problem	21/02/2023	21/02/2023
				A070504	Premature Discharge of Battery	21/02/2023	21/02/2023
		F1905	Device Revision or Replacement	A0705	Battery Problem	21/02/2023	21/02/2023
				A070504	Premature Discharge of Battery	21/02/2023	21/02/2023
CT321	Long-term surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem	21/02/2023	21/02/2023
				A070504	Premature Discharge of Battery	21/02/2023	21/02/2023
		F08	Hospitalization or Prolonged Hospitalization	A0705	Battery Problem	21/02/2023	21/02/2023
				A070504	Premature Discharge of Battery	21/02/2023	21/02/2023
		F1905	Device Revision or Replacement	A0705	Battery Problem	21/02/2023	21/02/2023
				A070504	Premature Discharge of Battery	21/02/2023	21/02/2023
CT449	Generators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem	21/02/2023	21/02/2023
				A070504	Premature Discharge of Battery	21/02/2023	21/02/2023
		F08	Hospitalization or Prolonged Hospitalization	A0705	Battery Problem	21/02/2023	21/02/2023
				A070504	Premature Discharge of Battery	21/02/2023	21/02/2023
		F1905	Device Revision or Replacement	A0705	Battery Problem	21/02/2023	21/02/2023
				A070504	Premature Discharge of Battery	21/02/2023	21/02/2023
CT981	Single use	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem	21/02/2023	21/02/2023
				A070504	Premature Discharge of Battery	21/02/2023	21/02/2023
		F08	Hospitalization or Prolonged Hospitalization	A0705	Battery Problem	21/02/2023	21/02/2023
				A070504	Premature Discharge of Battery	21/02/2023	21/02/2023
		F1905	Device Revision or Replacement	A0705	Battery Problem	21/02/2023	21/02/2023
				A070504	Premature Discharge of Battery	21/02/2023	21/02/2023

					CT983	Surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem	21/02/2023	21/02/2023
									A070504	Premature Discharge of Battery	21/02/2023	21/02/2023
							F08	Hospitalization or Prolonged Hospitalization	A0705	Battery Problem	21/02/2023	21/02/2023
									A070504	Premature Discharge of Battery	21/02/2023	21/02/2023
							F1905	Device Revision or Replacement	A0705	Battery Problem	21/02/2023	21/02/2023
									A070504	Premature Discharge of Battery	21/02/2023	21/02/2023
13872	2023	All other reportable incidents	47267	Single-chamber implantable pacemaker, rate-responsive	CT141	Electrodes/Leads and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
					CT145	Cardiovascular devices	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
					CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
					CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
					CT233	Surgical	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
					CT2406	Implantable	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
					CT2407	Active implantable	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
					CT2727	Active implant devices/ancillaries	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
					CT2750	Implantable cardiac electrical stimulators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
					CT2985	Active	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
					CT3001	Implantable metal	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
					CT3002	Implantable synthetic polymer	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
					CT3006	Implantable inorganic material	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
					CT314	Electrically powered	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
					CT321	Long-term surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
					CT449	Generators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
					CT981	Single use	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
					CT983	Surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
13873	2023	All other reportable incidents	35852	Single-chamber implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	15/07/2022	15/07/2022
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	15/07/2022	15/07/2022

							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	15/07/2022	15/07/2022
13874	2023	All other reportable incidents	47265	Dual-chamber implantable pacemaker, rate-responsive	CT141	Electrodes/Leads and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	06/06/2023	06/06/2023
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	06/06/2023	06/06/2023
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	06/06/2023	06/06/2023
					CT145	Cardiovascular devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	06/06/2023	06/06/2023
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	06/06/2023	06/06/2023
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	06/06/2023	06/06/2023
					CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	06/06/2023	06/06/2023
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	06/06/2023	06/06/2023
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	06/06/2023	06/06/2023
					CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	06/06/2023	06/06/2023
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	06/06/2023	06/06/2023
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	06/06/2023	06/06/2023
					CT233	Surgical	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	06/06/2023	06/06/2023
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	06/06/2023	06/06/2023
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	06/06/2023	06/06/2023
					CT2406	Implantable	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	06/06/2023	06/06/2023
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	06/06/2023	06/06/2023
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	06/06/2023	06/06/2023
					CT2407	Active implantable	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	06/06/2023	06/06/2023
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	06/06/2023	06/06/2023
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	06/06/2023	06/06/2023
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	06/06/2023	06/06/2023
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	06/06/2023	06/06/2023
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	06/06/2023	06/06/2023
					CT2727	Active implant devices/ancillaries	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	06/06/2023	06/06/2023
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	06/06/2023	06/06/2023
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	06/06/2023	06/06/2023
					CT2750	Implantable cardiac electrical stimulators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	06/06/2023	06/06/2023
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	06/06/2023	06/06/2023
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	06/06/2023	06/06/2023
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	06/06/2023	06/06/2023
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	06/06/2023	06/06/2023
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	06/06/2023	06/06/2023
					CT2985	Active	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	06/06/2023	06/06/2023
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	06/06/2023	06/06/2023
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	06/06/2023	06/06/2023
					CT3001	Implantable metal	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	06/06/2023	06/06/2023
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	06/06/2023	06/06/2023
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	06/06/2023	06/06/2023
					CT3002	Implantable synthetic polymer	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	06/06/2023	06/06/2023
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	06/06/2023	06/06/2023
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	06/06/2023	06/06/2023
					CT3006	Implantable inorganic material	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	06/06/2023	06/06/2023
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	06/06/2023	06/06/2023
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	06/06/2023	06/06/2023
					CT314	Electrically powered	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	06/06/2023	06/06/2023
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	06/06/2023	06/06/2023
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	06/06/2023	06/06/2023
					CT321	Long-term surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	06/06/2023	06/06/2023
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	06/06/2023	06/06/2023
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	06/06/2023	06/06/2023
					CT449	Generators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	06/06/2023	06/06/2023
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	06/06/2023	06/06/2023
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	06/06/2023	06/06/2023
					CT981	Single use	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	06/06/2023	06/06/2023
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	06/06/2023	06/06/2023
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	06/06/2023	06/06/2023
					CT983	Surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	06/06/2023	06/06/2023
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	06/06/2023	06/06/2023
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	06/06/2023	06/06/2023
13875	2023	All other reportable incidents	47265	Dual-chamber implantable pacemaker, rate-responsive	CT141	Electrodes/Leads and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A0908	Incorrect, Inadequate or Imprecise Result or Readings		
							F26	No Health Consequences or Impact	A0908	Incorrect, Inadequate or Imprecise Result or Readings		
					CT145	Cardiovascular devices	E2403	No Clinical Signs, Symptoms or Conditions	A0908	Incorrect, Inadequate or Imprecise Result or Readings		
							F26	No Health Consequences or Impact	A0908	Incorrect, Inadequate or Imprecise Result or Readings		
					CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A0908	Incorrect, Inadequate or Imprecise Result or Readings		
							F26	No Health Consequences or Impact	A0908	Incorrect, Inadequate or Imprecise Result or Readings		
					CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A0908	Incorrect, Inadequate or Imprecise Result or Readings		
							F26	No Health Consequences or Impact	A0908	Incorrect, Inadequate or Imprecise Result or Readings		
					CT233	Surgical	E2403	No Clinical Signs, Symptoms or Conditions	A0908	Incorrect, Inadequate or Imprecise Result or Readings		
							F26	No Health Consequences or Impact	A0908	Incorrect, Inadequate or Imprecise Result or Readings		
					CT2406	Implantable	E2403	No Clinical Signs, Symptoms or Conditions	A0908	Incorrect, Inadequate or Imprecise Result or Readings		
							F26	No Health Consequences or Impact	A0908	Incorrect, Inadequate or Imprecise Result or Readings		
					CT2407	Active implantable	E2403	No Clinical Signs, Symptoms or Conditions	A0908	Incorrect, Inadequate or Imprecise Result or Readings		
							F26	No Health Consequences or Impact	A0908	Incorrect, Inadequate or Imprecise Result or Readings		
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A0908	Incorrect, Inadequate or Imprecise Result or Readings		

							F26	No Health Consequences or Impact	A0908	Incorrect, Inadequate or Imprecise Result or Readings		
					CT2727	Active implant devices/ancillaries	E2403	No Clinical Signs, Symptoms or Conditions	A0908	Incorrect, Inadequate or Imprecise Result or Readings		
							F26	No Health Consequences or Impact	A0908	Incorrect, Inadequate or Imprecise Result or Readings		
					CT2750	Implantable cardiac electrical stimulators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A0908	Incorrect, Inadequate or Imprecise Result or Readings		
							F26	No Health Consequences or Impact	A0908	Incorrect, Inadequate or Imprecise Result or Readings		
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A0908	Incorrect, Inadequate or Imprecise Result or Readings		
							F26	No Health Consequences or Impact	A0908	Incorrect, Inadequate or Imprecise Result or Readings		
					CT2985	Active	E2403	No Clinical Signs, Symptoms or Conditions	A0908	Incorrect, Inadequate or Imprecise Result or Readings		
							F26	No Health Consequences or Impact	A0908	Incorrect, Inadequate or Imprecise Result or Readings		
					CT3001	Implantable metal	E2403	No Clinical Signs, Symptoms or Conditions	A0908	Incorrect, Inadequate or Imprecise Result or Readings		
							F26	No Health Consequences or Impact	A0908	Incorrect, Inadequate or Imprecise Result or Readings		
					CT3002	Implantable synthetic polymer	E2403	No Clinical Signs, Symptoms or Conditions	A0908	Incorrect, Inadequate or Imprecise Result or Readings		
							F26	No Health Consequences or Impact	A0908	Incorrect, Inadequate or Imprecise Result or Readings		
					CT3006	Implantable inorganic material	E2403	No Clinical Signs, Symptoms or Conditions	A0908	Incorrect, Inadequate or Imprecise Result or Readings		
							F26	No Health Consequences or Impact	A0908	Incorrect, Inadequate or Imprecise Result or Readings		
					CT314	Electrically powered	E2403	No Clinical Signs, Symptoms or Conditions	A0908	Incorrect, Inadequate or Imprecise Result or Readings		
							F26	No Health Consequences or Impact	A0908	Incorrect, Inadequate or Imprecise Result or Readings		
					CT321	Long-term surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A0908	Incorrect, Inadequate or Imprecise Result or Readings		
							F26	No Health Consequences or Impact	A0908	Incorrect, Inadequate or Imprecise Result or Readings		
					CT449	Generators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A0908	Incorrect, Inadequate or Imprecise Result or Readings		
							F26	No Health Consequences or Impact	A0908	Incorrect, Inadequate or Imprecise Result or Readings		
					CT981	Single use	E2403	No Clinical Signs, Symptoms or Conditions	A0908	Incorrect, Inadequate or Imprecise Result or Readings		
							F26	No Health Consequences or Impact	A0908	Incorrect, Inadequate or Imprecise Result or Readings		
					CT983	Surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A0908	Incorrect, Inadequate or Imprecise Result or Readings		
							F26	No Health Consequences or Impact	A0908	Incorrect, Inadequate or Imprecise Result or Readings		
13876	2023	All other reportable incidents	35852	Single-chamber implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A070909	Oversensing		
									A0712	Pacing Problem		
									A072201	High Impedance		
									A1301	Failure to Read Input Signal		
							F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A070909	Oversensing		
									A0712	Pacing Problem		
									A072201	High Impedance		
									A1301	Failure to Read Input Signal		
13877	2023	All other reportable incidents	47263	Cardiac resynchronization therapy implantable pacemaker	CT1685	Cardiac pacemakers	E0109	Convulsion/Seizure	A070909	Oversensing	01/11/2022	01/11/2022
									A0712	Pacing Problem	01/11/2022	01/11/2022
									A0908	Incorrect, Inadequate or Imprecise Result or Readings	01/11/2022	01/11/2022
							E011903	Syncope/Fainting	A070909	Oversensing	01/11/2022	01/11/2022
									A0712	Pacing Problem	01/11/2022	01/11/2022
									A0908	Incorrect, Inadequate or Imprecise Result or Readings	01/11/2022	01/11/2022
							E060101	Asystole	A070909	Oversensing	01/11/2022	01/11/2022
									A0712	Pacing Problem	01/11/2022	01/11/2022
									A0908	Incorrect, Inadequate or Imprecise Result or Readings	01/11/2022	01/11/2022
							F08	Hospitalization or Prolonged Hospitalization	A070909	Oversensing	01/11/2022	01/11/2022
									A0712	Pacing Problem	01/11/2022	01/11/2022
									A0908	Incorrect, Inadequate or Imprecise Result or Readings	01/11/2022	01/11/2022
							F1001	Absence of Treatment	A070909	Oversensing	01/11/2022	01/11/2022
									A0712	Pacing Problem	01/11/2022	01/11/2022
									A0908	Incorrect, Inadequate or Imprecise Result or Readings	01/11/2022	01/11/2022
							F1905	Device Revision or Replacement	A070909	Oversensing	01/11/2022	01/11/2022
									A0712	Pacing Problem	01/11/2022	01/11/2022
									A0908	Incorrect, Inadequate or Imprecise Result or Readings	01/11/2022	01/11/2022
					CT257	Implantable cardiac pacemakers	E0109	Convulsion/Seizure	A070909	Oversensing	01/11/2022	01/11/2022
									A0712	Pacing Problem	01/11/2022	01/11/2022
									A0908	Incorrect, Inadequate or Imprecise Result or Readings	01/11/2022	01/11/2022
							E011903	Syncope/Fainting	A070909	Oversensing	01/11/2022	01/11/2022
									A0712	Pacing Problem	01/11/2022	01/11/2022
									A0908	Incorrect, Inadequate or Imprecise Result or Readings	01/11/2022	01/11/2022
							E060101	Asystole	A070909	Oversensing	01/11/2022	01/11/2022
									A0712	Pacing Problem	01/11/2022	01/11/2022
									A0908	Incorrect, Inadequate or Imprecise Result or Readings	01/11/2022	01/11/2022
							F08	Hospitalization or Prolonged Hospitalization	A070909	Oversensing	01/11/2022	01/11/2022
									A0712	Pacing Problem	01/11/2022	01/11/2022

								A0908	Incorrect, Inadequate or Imprecise Result or Readings	01/11/2022	01/11/2022
						F1001	Absence of Treatment	A070909	Oversensing	01/11/2022	01/11/2022
								A0712	Pacing Problem	01/11/2022	01/11/2022
								A0908	Incorrect, Inadequate or Imprecise Result or Readings	01/11/2022	01/11/2022
						F1905	Device Revision or Replacement	A070909	Oversensing	01/11/2022	01/11/2022
								A0712	Pacing Problem	01/11/2022	01/11/2022
								A0908	Incorrect, Inadequate or Imprecise Result or Readings	01/11/2022	01/11/2022
				CT2866	Implantable cardiac electrical stimulators	E0109	Convulsion/Seizure	A070909	Oversensing	01/11/2022	01/11/2022
								A0712	Pacing Problem	01/11/2022	01/11/2022
								A0908	Incorrect, Inadequate or Imprecise Result or Readings	01/11/2022	01/11/2022
						E011903	Syncope/Fainting	A070909	Oversensing	01/11/2022	01/11/2022
								A0712	Pacing Problem	01/11/2022	01/11/2022
								A0908	Incorrect, Inadequate or Imprecise Result or Readings	01/11/2022	01/11/2022
						E060101	Asystole	A070909	Oversensing	01/11/2022	01/11/2022
								A0712	Pacing Problem	01/11/2022	01/11/2022
								A0908	Incorrect, Inadequate or Imprecise Result or Readings	01/11/2022	01/11/2022
						F08	Hospitalization or Prolonged Hospitalization	A070909	Oversensing	01/11/2022	01/11/2022
								A0712	Pacing Problem	01/11/2022	01/11/2022
								A0908	Incorrect, Inadequate or Imprecise Result or Readings	01/11/2022	01/11/2022
						F1001	Absence of Treatment	A070909	Oversensing	01/11/2022	01/11/2022
								A0712	Pacing Problem	01/11/2022	01/11/2022
								A0908	Incorrect, Inadequate or Imprecise Result or Readings	01/11/2022	01/11/2022
						F1905	Device Revision or Replacement	A070909	Oversensing	01/11/2022	01/11/2022
								A0712	Pacing Problem	01/11/2022	01/11/2022
								A0908	Incorrect, Inadequate or Imprecise Result or Readings	01/11/2022	01/11/2022
13878	2023	All other reportable incidents	47265	Dual-chamber implantable pacemaker, rate-responsive	CT141	Electrodes/Leads and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
							F24	Insufficient Health Impact Information	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
					CT145	Cardiovascular devices	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
							F24	Insufficient Health Impact Information	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
					CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
							F24	Insufficient Health Impact Information	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
					CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
							F24	Insufficient Health Impact Information	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
					CT233	Surgical	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
							F24	Insufficient Health Impact Information	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
					CT2406	Implantable	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
							F24	Insufficient Health Impact Information	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
					CT2407	Active implantable	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
							F24	Insufficient Health Impact Information	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
							F24	Insufficient Health Impact Information	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
					CT2727	Active implant devices/ancillaries	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
							F24	Insufficient Health Impact Information	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
					CT2750	Implantable cardiac electrical stimulators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
							F24	Insufficient Health Impact Information	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
							F24	Insufficient Health Impact Information	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
					CT2985	Active	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
							F24	Insufficient Health Impact Information	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
					CT3001	Implantable metal	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
							F24	Insufficient Health Impact Information	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
					CT3002	Implantable synthetic polymer	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
							F24	Insufficient Health Impact Information	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		

					CT3006	Implantable inorganic material	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
							F24	Insufficient Health Impact Information	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
					CT314	Electrically powered	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
							F24	Insufficient Health Impact Information	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
					CT321	Long-term surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
							F24	Insufficient Health Impact Information	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
					CT449	Generators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
							F24	Insufficient Health Impact Information	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
					CT981	Single use	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
							F24	Insufficient Health Impact Information	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
					CT983	Surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
							F24	Insufficient Health Impact Information	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
13879	2023	All other reportable incidents	47263	Cardiac resynchronization therapy implantable pacemaker	CT1684	Cardiac pacemakers and associated devices	E011903	Syncope/Fainting	A0711	Interrogation Problem	28/04/2023	28/04/2023
									A0712	Pacing Problem	28/04/2023	28/04/2023
							F1001	Absence of Treatment	A0711	Interrogation Problem	28/04/2023	28/04/2023
									A0712	Pacing Problem	28/04/2023	28/04/2023
					CT1685	Cardiac pacemakers	E011903	Syncope/Fainting	A0711	Interrogation Problem	28/04/2023	28/04/2023
									A0712	Pacing Problem	28/04/2023	28/04/2023
							F1001	Absence of Treatment	A0711	Interrogation Problem	28/04/2023	28/04/2023
									A0712	Pacing Problem	28/04/2023	28/04/2023
					CT257	Implantable cardiac pacemakers	E011903	Syncope/Fainting	A0711	Interrogation Problem	28/04/2023	28/04/2023
									A0712	Pacing Problem	28/04/2023	28/04/2023
							F1001	Absence of Treatment	A0711	Interrogation Problem	28/04/2023	28/04/2023
									A0712	Pacing Problem	28/04/2023	28/04/2023
					CT2866	Implantable cardiac electrical stimulators	E011903	Syncope/Fainting	A0711	Interrogation Problem	28/04/2023	28/04/2023
									A0712	Pacing Problem	28/04/2023	28/04/2023
							F1001	Absence of Treatment	A0711	Interrogation Problem	28/04/2023	28/04/2023
									A0712	Pacing Problem	28/04/2023	28/04/2023
13880	2023	All other reportable incidents	47265	Dual-chamber implantable pacemaker, rate-responsive	CT141	Electrodes/Leads and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem	04/07/2023	04/07/2023
									A070504	Premature Discharge of Battery	04/07/2023	04/07/2023
							F08	Hospitalization or Prolonged Hospitalization	A0705	Battery Problem	04/07/2023	04/07/2023
									A070504	Premature Discharge of Battery	04/07/2023	04/07/2023
							F1905	Device Revision or Replacement	A0705	Battery Problem	04/07/2023	04/07/2023
									A070504	Premature Discharge of Battery	04/07/2023	04/07/2023
					CT145	Cardiovascular devices	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem	04/07/2023	04/07/2023
									A070504	Premature Discharge of Battery	04/07/2023	04/07/2023
							F08	Hospitalization or Prolonged Hospitalization	A0705	Battery Problem	04/07/2023	04/07/2023
									A070504	Premature Discharge of Battery	04/07/2023	04/07/2023
							F1905	Device Revision or Replacement	A0705	Battery Problem	04/07/2023	04/07/2023
									A070504	Premature Discharge of Battery	04/07/2023	04/07/2023
					CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem	04/07/2023	04/07/2023
									A070504	Premature Discharge of Battery	04/07/2023	04/07/2023
							F08	Hospitalization or Prolonged Hospitalization	A0705	Battery Problem	04/07/2023	04/07/2023
									A070504	Premature Discharge of Battery	04/07/2023	04/07/2023
							F1905	Device Revision or Replacement	A0705	Battery Problem	04/07/2023	04/07/2023
									A070504	Premature Discharge of Battery	04/07/2023	04/07/2023
					CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem	04/07/2023	04/07/2023
									A070504	Premature Discharge of Battery	04/07/2023	04/07/2023
							F08	Hospitalization or Prolonged Hospitalization	A0705	Battery Problem	04/07/2023	04/07/2023
									A070504	Premature Discharge of Battery	04/07/2023	04/07/2023
							F1905	Device Revision or Replacement	A0705	Battery Problem	04/07/2023	04/07/2023
									A070504	Premature Discharge of Battery	04/07/2023	04/07/2023
					CT233	Surgical	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem	04/07/2023	04/07/2023
									A070504	Premature Discharge of Battery	04/07/2023	04/07/2023
							F08	Hospitalization or Prolonged Hospitalization	A0705	Battery Problem	04/07/2023	04/07/2023
									A070504	Premature Discharge of Battery	04/07/2023	04/07/2023
							F1905	Device Revision or Replacement	A0705	Battery Problem	04/07/2023	04/07/2023
									A070504	Premature Discharge of Battery	04/07/2023	04/07/2023
					CT2406	Implantable	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem	04/07/2023	04/07/2023
									A070504	Premature Discharge of Battery	04/07/2023	04/07/2023
							F08	Hospitalization or Prolonged Hospitalization	A0705	Battery Problem	04/07/2023	04/07/2023
									A070504	Premature Discharge of Battery	04/07/2023	04/07/2023
							F1905	Device Revision or Replacement	A0705	Battery Problem	04/07/2023	04/07/2023
									A070504	Premature Discharge of Battery	04/07/2023	04/07/2023
					CT2407	Active implantable	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem	04/07/2023	04/07/2023
									A070504	Premature Discharge of Battery	04/07/2023	04/07/2023
							F08	Hospitalization or Prolonged Hospitalization	A0705	Battery Problem	04/07/2023	04/07/2023
									A070504	Premature Discharge of Battery	04/07/2023	04/07/2023
							F1905	Device Revision or Replacement	A0705	Battery Problem	04/07/2023	04/07/2023
									A070504	Premature Discharge of Battery	04/07/2023	04/07/2023
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem	04/07/2023	04/07/2023
									A070504	Premature Discharge of Battery	04/07/2023	04/07/2023
							F08	Hospitalization or Prolonged Hospitalization	A0705	Battery Problem	04/07/2023	04/07/2023
									A070504	Premature Discharge of Battery	04/07/2023	04/07/2023
							F1905	Device Revision or Replacement	A0705	Battery Problem	04/07/2023	04/07/2023
									A070504	Premature Discharge of Battery	04/07/2023	04/07/2023

					CT2727	Active implant devices/ancillaries	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem	04/07/2023	04/07/2023
									A070504	Premature Discharge of Battery	04/07/2023	04/07/2023
							F08	Hospitalization or Prolonged Hospitalization	A0705	Battery Problem	04/07/2023	04/07/2023
									A070504	Premature Discharge of Battery	04/07/2023	04/07/2023
							F1905	Device Revision or Replacement	A0705	Battery Problem	04/07/2023	04/07/2023
									A070504	Premature Discharge of Battery	04/07/2023	04/07/2023
					CT2750	Implantable cardiac electrical stimulators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem	04/07/2023	04/07/2023
									A070504	Premature Discharge of Battery	04/07/2023	04/07/2023
							F08	Hospitalization or Prolonged Hospitalization	A0705	Battery Problem	04/07/2023	04/07/2023
									A070504	Premature Discharge of Battery	04/07/2023	04/07/2023
							F1905	Device Revision or Replacement	A0705	Battery Problem	04/07/2023	04/07/2023
									A070504	Premature Discharge of Battery	04/07/2023	04/07/2023
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem	04/07/2023	04/07/2023
									A070504	Premature Discharge of Battery	04/07/2023	04/07/2023
							F08	Hospitalization or Prolonged Hospitalization	A0705	Battery Problem	04/07/2023	04/07/2023
									A070504	Premature Discharge of Battery	04/07/2023	04/07/2023
							F1905	Device Revision or Replacement	A0705	Battery Problem	04/07/2023	04/07/2023
									A070504	Premature Discharge of Battery	04/07/2023	04/07/2023
					CT2985	Active	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem	04/07/2023	04/07/2023
									A070504	Premature Discharge of Battery	04/07/2023	04/07/2023
							F08	Hospitalization or Prolonged Hospitalization	A0705	Battery Problem	04/07/2023	04/07/2023
									A070504	Premature Discharge of Battery	04/07/2023	04/07/2023
							F1905	Device Revision or Replacement	A0705	Battery Problem	04/07/2023	04/07/2023
									A070504	Premature Discharge of Battery	04/07/2023	04/07/2023
					CT3001	Implantable metal	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem	04/07/2023	04/07/2023
									A070504	Premature Discharge of Battery	04/07/2023	04/07/2023
							F08	Hospitalization or Prolonged Hospitalization	A0705	Battery Problem	04/07/2023	04/07/2023
									A070504	Premature Discharge of Battery	04/07/2023	04/07/2023
							F1905	Device Revision or Replacement	A0705	Battery Problem	04/07/2023	04/07/2023
									A070504	Premature Discharge of Battery	04/07/2023	04/07/2023
					CT3002	Implantable synthetic polymer	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem	04/07/2023	04/07/2023
									A070504	Premature Discharge of Battery	04/07/2023	04/07/2023
							F08	Hospitalization or Prolonged Hospitalization	A0705	Battery Problem	04/07/2023	04/07/2023
									A070504	Premature Discharge of Battery	04/07/2023	04/07/2023
							F1905	Device Revision or Replacement	A0705	Battery Problem	04/07/2023	04/07/2023
									A070504	Premature Discharge of Battery	04/07/2023	04/07/2023
					CT3006	Implantable inorganic material	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem	04/07/2023	04/07/2023
									A070504	Premature Discharge of Battery	04/07/2023	04/07/2023
							F08	Hospitalization or Prolonged Hospitalization	A0705	Battery Problem	04/07/2023	04/07/2023
									A070504	Premature Discharge of Battery	04/07/2023	04/07/2023
							F1905	Device Revision or Replacement	A0705	Battery Problem	04/07/2023	04/07/2023
									A070504	Premature Discharge of Battery	04/07/2023	04/07/2023
					CT314	Electrically powered	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem	04/07/2023	04/07/2023
									A070504	Premature Discharge of Battery	04/07/2023	04/07/2023
							F08	Hospitalization or Prolonged Hospitalization	A0705	Battery Problem	04/07/2023	04/07/2023
									A070504	Premature Discharge of Battery	04/07/2023	04/07/2023
							F1905	Device Revision or Replacement	A0705	Battery Problem	04/07/2023	04/07/2023
									A070504	Premature Discharge of Battery	04/07/2023	04/07/2023
					CT321	Long-term surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem	04/07/2023	04/07/2023
									A070504	Premature Discharge of Battery	04/07/2023	04/07/2023
							F08	Hospitalization or Prolonged Hospitalization	A0705	Battery Problem	04/07/2023	04/07/2023
									A070504	Premature Discharge of Battery	04/07/2023	04/07/2023
							F1905	Device Revision or Replacement	A0705	Battery Problem	04/07/2023	04/07/2023
									A070504	Premature Discharge of Battery	04/07/2023	04/07/2023
					CT449	Generators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem	04/07/2023	04/07/2023
									A070504	Premature Discharge of Battery	04/07/2023	04/07/2023
							F08	Hospitalization or Prolonged Hospitalization	A0705	Battery Problem	04/07/2023	04/07/2023
									A070504	Premature Discharge of Battery	04/07/2023	04/07/2023
							F1905	Device Revision or Replacement	A0705	Battery Problem	04/07/2023	04/07/2023
									A070504	Premature Discharge of Battery	04/07/2023	04/07/2023
					CT981	Single use	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem	04/07/2023	04/07/2023
									A070504	Premature Discharge of Battery	04/07/2023	04/07/2023
							F08	Hospitalization or Prolonged Hospitalization	A0705	Battery Problem	04/07/2023	04/07/2023
									A070504	Premature Discharge of Battery	04/07/2023	04/07/2023
							F1905	Device Revision or Replacement	A0705	Battery Problem	04/07/2023	04/07/2023
									A070504	Premature Discharge of Battery	04/07/2023	04/07/2023
					CT983	Surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem	04/07/2023	04/07/2023
									A070504	Premature Discharge of Battery	04/07/2023	04/07/2023
							F08	Hospitalization or Prolonged Hospitalization	A0705	Battery Problem	04/07/2023	04/07/2023
									A070504	Premature Discharge of Battery	04/07/2023	04/07/2023
							F1905	Device Revision or Replacement	A0705	Battery Problem	04/07/2023	04/07/2023
									A070504	Premature Discharge of Battery	04/07/2023	04/07/2023
13881	2023	All other reportable incidents	36241	Pacemaker/defibrillator lead	CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A072201	High Impedance		
							F26	No Health Consequences or Impact	A072201	High Impedance		
13882	2023	All other reportable incidents	47265	Dual-chamber implantable pacemaker, rate-responsive	CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem	16/08/2022	16/08/2022
									A070504	Premature Discharge of Battery	16/08/2022	16/08/2022
							F08	Hospitalization or Prolonged Hospitalization	A0705	Battery Problem	16/08/2022	16/08/2022
									A070504	Premature Discharge of Battery	16/08/2022	16/08/2022
							F1905	Device Revision or Replacement	A0705	Battery Problem	16/08/2022	16/08/2022
									A070504	Premature Discharge of Battery	16/08/2022	16/08/2022
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem	16/08/2022	16/08/2022
									A070504	Premature Discharge of Battery	16/08/2022	16/08/2022
							F08	Hospitalization or Prolonged Hospitalization	A0705	Battery Problem	16/08/2022	16/08/2022
									A070504	Premature Discharge of Battery	16/08/2022	16/08/2022
							F1905	Device Revision or Replacement	A0705	Battery Problem	16/08/2022	16/08/2022
									A070504	Premature Discharge of Battery	16/08/2022	16/08/2022
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem	16/08/2022	16/08/2022
									A070504	Premature Discharge of Battery	16/08/2022	16/08/2022

							F08	Hospitalization or Prolonged Hospitalization	A0705	Battery Problem	16/08/2022	16/08/2022
									A070504	Premature Discharge of Battery	16/08/2022	16/08/2022
							F1905	Device Revision or Replacement	A0705	Battery Problem	16/08/2022	16/08/2022
									A070504	Premature Discharge of Battery	16/08/2022	16/08/2022
13883	2023	All other reportable incidents	47205	Cardiac pulse generator programmer	CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A150105	Key or Button Unresponsive/not Working		
							F26	No Health Consequences or Impact	A150105	Key or Button Unresponsive/not Working		
13884	2023	All other reportable incidents	36241	Pacemaker/defibrillator lead	CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A072201	High Impedance		
							F26	No Health Consequences or Impact	A072201	High Impedance		
13885	2023	All other reportable incidents	47265	Dual-chamber implantable pacemaker, rate-responsive	CT141	Electrodes/Leads and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	28/06/2023	28/06/2023
									A0908	Incorrect, Inadequate or Imprecise Result or Readings	28/06/2023	28/06/2023
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	28/06/2023	28/06/2023
									A0908	Incorrect, Inadequate or Imprecise Result or Readings	28/06/2023	28/06/2023
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	28/06/2023	28/06/2023
									A0908	Incorrect, Inadequate or Imprecise Result or Readings	28/06/2023	28/06/2023
					CT145	Cardiovascular devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	28/06/2023	28/06/2023
									A0908	Incorrect, Inadequate or Imprecise Result or Readings	28/06/2023	28/06/2023
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	28/06/2023	28/06/2023
									A0908	Incorrect, Inadequate or Imprecise Result or Readings	28/06/2023	28/06/2023
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	28/06/2023	28/06/2023
									A0908	Incorrect, Inadequate or Imprecise Result or Readings	28/06/2023	28/06/2023
					CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	28/06/2023	28/06/2023
									A0908	Incorrect, Inadequate or Imprecise Result or Readings	28/06/2023	28/06/2023
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	28/06/2023	28/06/2023
									A0908	Incorrect, Inadequate or Imprecise Result or Readings	28/06/2023	28/06/2023
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	28/06/2023	28/06/2023
									A0908	Incorrect, Inadequate or Imprecise Result or Readings	28/06/2023	28/06/2023
					CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	28/06/2023	28/06/2023
									A0908	Incorrect, Inadequate or Imprecise Result or Readings	28/06/2023	28/06/2023
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	28/06/2023	28/06/2023
									A0908	Incorrect, Inadequate or Imprecise Result or Readings	28/06/2023	28/06/2023
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	28/06/2023	28/06/2023
									A0908	Incorrect, Inadequate or Imprecise Result or Readings	28/06/2023	28/06/2023
					CT233	Surgical	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	28/06/2023	28/06/2023
									A0908	Incorrect, Inadequate or Imprecise Result or Readings	28/06/2023	28/06/2023
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	28/06/2023	28/06/2023
									A0908	Incorrect, Inadequate or Imprecise Result or Readings	28/06/2023	28/06/2023
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	28/06/2023	28/06/2023
									A0908	Incorrect, Inadequate or Imprecise Result or Readings	28/06/2023	28/06/2023
					CT2406	Implantable	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	28/06/2023	28/06/2023
									A0908	Incorrect, Inadequate or Imprecise Result or Readings	28/06/2023	28/06/2023
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	28/06/2023	28/06/2023
									A0908	Incorrect, Inadequate or Imprecise Result or Readings	28/06/2023	28/06/2023
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	28/06/2023	28/06/2023
									A0908	Incorrect, Inadequate or Imprecise Result or Readings	28/06/2023	28/06/2023
					CT2407	Active implantable	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	28/06/2023	28/06/2023
									A0908	Incorrect, Inadequate or Imprecise Result or Readings	28/06/2023	28/06/2023
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	28/06/2023	28/06/2023
									A0908	Incorrect, Inadequate or Imprecise Result or Readings	28/06/2023	28/06/2023
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	28/06/2023	28/06/2023
									A0908	Incorrect, Inadequate or Imprecise Result or Readings	28/06/2023	28/06/2023
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	28/06/2023	28/06/2023
									A0908	Incorrect, Inadequate or Imprecise Result or Readings	28/06/2023	28/06/2023
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	28/06/2023	28/06/2023
									A0908	Incorrect, Inadequate or Imprecise Result or Readings	28/06/2023	28/06/2023
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	28/06/2023	28/06/2023
									A0908	Incorrect, Inadequate or Imprecise Result or Readings	28/06/2023	28/06/2023
					CT2727	Active implant devices/ancillaries	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	28/06/2023	28/06/2023
									A0908	Incorrect, Inadequate or Imprecise Result or Readings	28/06/2023	28/06/2023
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	28/06/2023	28/06/2023
									A0908	Incorrect, Inadequate or Imprecise Result or Readings	28/06/2023	28/06/2023
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	28/06/2023	28/06/2023
									A0908	Incorrect, Inadequate or Imprecise Result or Readings	28/06/2023	28/06/2023
					CT2750	Implantable cardiac electrical stimulators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	28/06/2023	28/06/2023
									A0908	Incorrect, Inadequate or Imprecise Result or Readings	28/06/2023	28/06/2023
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	28/06/2023	28/06/2023
									A0908	Incorrect, Inadequate or Imprecise Result or Readings	28/06/2023	28/06/2023

							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	28/06/2023	28/06/2023
									A0908	Incorrect, Inadequate or Imprecise Result or Readings	28/06/2023	28/06/2023
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	28/06/2023	28/06/2023
									A0908	Incorrect, Inadequate or Imprecise Result or Readings	28/06/2023	28/06/2023
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	28/06/2023	28/06/2023
									A0908	Incorrect, Inadequate or Imprecise Result or Readings	28/06/2023	28/06/2023
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	28/06/2023	28/06/2023
									A0908	Incorrect, Inadequate or Imprecise Result or Readings	28/06/2023	28/06/2023
					CT2985	Active	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	28/06/2023	28/06/2023
									A0908	Incorrect, Inadequate or Imprecise Result or Readings	28/06/2023	28/06/2023
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	28/06/2023	28/06/2023
									A0908	Incorrect, Inadequate or Imprecise Result or Readings	28/06/2023	28/06/2023
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	28/06/2023	28/06/2023
									A0908	Incorrect, Inadequate or Imprecise Result or Readings	28/06/2023	28/06/2023
					CT3001	Implantable metal	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	28/06/2023	28/06/2023
									A0908	Incorrect, Inadequate or Imprecise Result or Readings	28/06/2023	28/06/2023
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	28/06/2023	28/06/2023
									A0908	Incorrect, Inadequate or Imprecise Result or Readings	28/06/2023	28/06/2023
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	28/06/2023	28/06/2023
									A0908	Incorrect, Inadequate or Imprecise Result or Readings	28/06/2023	28/06/2023
					CT3002	Implantable synthetic polymer	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	28/06/2023	28/06/2023
									A0908	Incorrect, Inadequate or Imprecise Result or Readings	28/06/2023	28/06/2023
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	28/06/2023	28/06/2023
									A0908	Incorrect, Inadequate or Imprecise Result or Readings	28/06/2023	28/06/2023
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	28/06/2023	28/06/2023
									A0908	Incorrect, Inadequate or Imprecise Result or Readings	28/06/2023	28/06/2023
					CT3006	Implantable inorganic material	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	28/06/2023	28/06/2023
									A0908	Incorrect, Inadequate or Imprecise Result or Readings	28/06/2023	28/06/2023
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	28/06/2023	28/06/2023
									A0908	Incorrect, Inadequate or Imprecise Result or Readings	28/06/2023	28/06/2023
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	28/06/2023	28/06/2023
									A0908	Incorrect, Inadequate or Imprecise Result or Readings	28/06/2023	28/06/2023
					CT314	Electrically powered	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	28/06/2023	28/06/2023
									A0908	Incorrect, Inadequate or Imprecise Result or Readings	28/06/2023	28/06/2023
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	28/06/2023	28/06/2023
									A0908	Incorrect, Inadequate or Imprecise Result or Readings	28/06/2023	28/06/2023
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	28/06/2023	28/06/2023
									A0908	Incorrect, Inadequate or Imprecise Result or Readings	28/06/2023	28/06/2023
					CT321	Long-term surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	28/06/2023	28/06/2023
									A0908	Incorrect, Inadequate or Imprecise Result or Readings	28/06/2023	28/06/2023
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	28/06/2023	28/06/2023
									A0908	Incorrect, Inadequate or Imprecise Result or Readings	28/06/2023	28/06/2023
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	28/06/2023	28/06/2023
									A0908	Incorrect, Inadequate or Imprecise Result or Readings	28/06/2023	28/06/2023
					CT449	Generators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	28/06/2023	28/06/2023
									A0908	Incorrect, Inadequate or Imprecise Result or Readings	28/06/2023	28/06/2023
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	28/06/2023	28/06/2023
									A0908	Incorrect, Inadequate or Imprecise Result or Readings	28/06/2023	28/06/2023
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	28/06/2023	28/06/2023
									A0908	Incorrect, Inadequate or Imprecise Result or Readings	28/06/2023	28/06/2023
					CT981	Single use	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	28/06/2023	28/06/2023
									A0908	Incorrect, Inadequate or Imprecise Result or Readings	28/06/2023	28/06/2023
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	28/06/2023	28/06/2023
									A0908	Incorrect, Inadequate or Imprecise Result or Readings	28/06/2023	28/06/2023
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	28/06/2023	28/06/2023
									A0908	Incorrect, Inadequate or Imprecise Result or Readings	28/06/2023	28/06/2023
					CT983	Surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	28/06/2023	28/06/2023
									A0908	Incorrect, Inadequate or Imprecise Result or Readings	28/06/2023	28/06/2023
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	28/06/2023	28/06/2023
									A0908	Incorrect, Inadequate or Imprecise Result or Readings	28/06/2023	28/06/2023
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	28/06/2023	28/06/2023
									A0908	Incorrect, Inadequate or Imprecise Result or Readings	28/06/2023	28/06/2023
13886	2023	All other reportable incidents	47265	Dual-chamber implantable pacemaker, rate-responsive	CT141	Electrodes/Leads and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		

					CT145	Cardiovascular devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT233	Surgical	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT2406	Implantable	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT2407	Active implantable	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT2727	Active implant devices/ancillaries	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT2750	Implantable cardiac electrical stimulators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT2985	Active	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT3001	Implantable metal	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT3002	Implantable synthetic polymer	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT3006	Implantable inorganic material	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT314	Electrically powered	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT321	Long-term surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT449	Generators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT981	Single use	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT983	Surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
13887	2023		47265	Dual-chamber implantable pacemaker, rate-responsive	CT141	Electrodes/Leads and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	17/03/2023	17/03/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	17/03/2023	17/03/2023
					CT145	Cardiovascular devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	17/03/2023	17/03/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	17/03/2023	17/03/2023
					CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	17/03/2023	17/03/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	17/03/2023	17/03/2023
					CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	17/03/2023	17/03/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	17/03/2023	17/03/2023
					CT233	Surgical	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	17/03/2023	17/03/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	17/03/2023	17/03/2023
					CT2406	Implantable	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	17/03/2023	17/03/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	17/03/2023	17/03/2023
					CT2407	Active implantable	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	17/03/2023	17/03/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	17/03/2023	17/03/2023
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	17/03/2023	17/03/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	17/03/2023	17/03/2023
					CT2727	Active implant devices/ancillaries	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	17/03/2023	17/03/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	17/03/2023	17/03/2023
					CT2750	Implantable cardiac electrical stimulators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	17/03/2023	17/03/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	17/03/2023	17/03/2023
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	17/03/2023	17/03/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	17/03/2023	17/03/2023
					CT2985	Active	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	17/03/2023	17/03/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	17/03/2023	17/03/2023
					CT3001	Implantable metal	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	17/03/2023	17/03/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	17/03/2023	17/03/2023
					CT3002	Implantable synthetic polymer	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	17/03/2023	17/03/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	17/03/2023	17/03/2023
					CT3006	Implantable inorganic material	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	17/03/2023	17/03/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	17/03/2023	17/03/2023
					CT314	Electrically powered	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	17/03/2023	17/03/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	17/03/2023	17/03/2023
					CT321	Long-term surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	17/03/2023	17/03/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	17/03/2023	17/03/2023
					CT449	Generators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	17/03/2023	17/03/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	17/03/2023	17/03/2023
					CT981	Single use	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	17/03/2023	17/03/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	17/03/2023	17/03/2023
					CT983	Surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	17/03/2023	17/03/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	17/03/2023	17/03/2023
13888	2023		47265	Dual-chamber implantable pacemaker, rate-responsive	CT141	Electrodes/Leads and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	03/07/2023	03/07/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	03/07/2023	03/07/2023
					CT145	Cardiovascular devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	03/07/2023	03/07/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	03/07/2023	03/07/2023
					CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	03/07/2023	03/07/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	03/07/2023	03/07/2023
					CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	03/07/2023	03/07/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	03/07/2023	03/07/2023
					CT233	Surgical	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	03/07/2023	03/07/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	03/07/2023	03/07/2023

					CT2406	Implantable	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	03/07/2023	03/07/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	03/07/2023	03/07/2023
					CT2407	Active implantable	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	03/07/2023	03/07/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	03/07/2023	03/07/2023
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	03/07/2023	03/07/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	03/07/2023	03/07/2023
					CT2727	Active implant devices/ancillaries	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	03/07/2023	03/07/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	03/07/2023	03/07/2023
					CT2750	Implantable cardiac electrical stimulators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	03/07/2023	03/07/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	03/07/2023	03/07/2023
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	03/07/2023	03/07/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	03/07/2023	03/07/2023
					CT2985	Active	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	03/07/2023	03/07/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	03/07/2023	03/07/2023
					CT3001	Implantable metal	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	03/07/2023	03/07/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	03/07/2023	03/07/2023
					CT3002	Implantable synthetic polymer	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	03/07/2023	03/07/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	03/07/2023	03/07/2023
					CT3006	Implantable inorganic material	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	03/07/2023	03/07/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	03/07/2023	03/07/2023
					CT314	Electrically powered	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	03/07/2023	03/07/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	03/07/2023	03/07/2023
					CT321	Long-term surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	03/07/2023	03/07/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	03/07/2023	03/07/2023
					CT449	Generators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	03/07/2023	03/07/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	03/07/2023	03/07/2023
					CT981	Single use	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	03/07/2023	03/07/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	03/07/2023	03/07/2023
					CT983	Surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	03/07/2023	03/07/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	03/07/2023	03/07/2023
13889	2023		47265	Dual-chamber implantable pacemaker, rate-responsive	CT141	Electrodes/Leads and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	05/05/2023	05/05/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	05/05/2023	05/05/2023
					CT145	Cardiovascular devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	05/05/2023	05/05/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	05/05/2023	05/05/2023
					CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	05/05/2023	05/05/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	05/05/2023	05/05/2023
					CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	05/05/2023	05/05/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	05/05/2023	05/05/2023
					CT233	Surgical	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	05/05/2023	05/05/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	05/05/2023	05/05/2023
					CT2406	Implantable	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	05/05/2023	05/05/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	05/05/2023	05/05/2023
					CT2407	Active implantable	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	05/05/2023	05/05/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	05/05/2023	05/05/2023
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	05/05/2023	05/05/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	05/05/2023	05/05/2023
					CT2727	Active implant devices/ancillaries	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	05/05/2023	05/05/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	05/05/2023	05/05/2023
					CT2750	Implantable cardiac electrical stimulators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	05/05/2023	05/05/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	05/05/2023	05/05/2023
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	05/05/2023	05/05/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	05/05/2023	05/05/2023
					CT2985	Active	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	05/05/2023	05/05/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	05/05/2023	05/05/2023
					CT3001	Implantable metal	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	05/05/2023	05/05/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	05/05/2023	05/05/2023
					CT3002	Implantable synthetic polymer	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	05/05/2023	05/05/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	05/05/2023	05/05/2023
					CT3006	Implantable inorganic material	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	05/05/2023	05/05/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	05/05/2023	05/05/2023
					CT314	Electrically powered	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	05/05/2023	05/05/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	05/05/2023	05/05/2023
					CT321	Long-term surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	05/05/2023	05/05/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	05/05/2023	05/05/2023
					CT449	Generators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	05/05/2023	05/05/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	05/05/2023	05/05/2023
					CT981	Single use	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	05/05/2023	05/05/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	05/05/2023	05/05/2023
					CT983	Surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	05/05/2023	05/05/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	05/05/2023	05/05/2023
13890	2023	All other reportable incidents	63188	Left-ventricular wireless pacing system sensor/transmitter	CT145	Cardiovascular devices	F24	Insufficient Health Impact Information	A070801	Complete Loss of Power		
					CT1577	Wireless monitoring system transmitters	F24	Insufficient Health Impact Information	A070801	Complete Loss of Power		
					CT1684	Cardiac pacemakers and associated devices	F24	Insufficient Health Impact Information	A070801	Complete Loss of Power		
					CT2727	Active implant devices/ancillaries	F24	Insufficient Health Impact Information	A070801	Complete Loss of Power		
					CT2750	Implantable cardiac electrical stimulators and associated devices	F24	Insufficient Health Impact Information	A070801	Complete Loss of Power		
					CT2885	Physiological monitoring/measuring system transmitters/receivers/transceivers	F24	Insufficient Health Impact Information	A070801	Complete Loss of Power		
13891	2023	All other reportable incidents	63188	Left-ventricular wireless pacing system sensor/transmitter	CT145	Cardiovascular devices			A0701	Capturing Problem		
					CT1577	Wireless monitoring system transmitters			A0701	Capturing Problem		
					CT1684	Cardiac pacemakers and associated devices			A0701	Capturing Problem		
					CT2727	Active implant devices/ancillaries			A0701	Capturing Problem		
					CT2750	Implantable cardiac electrical stimulators and associated devices			A0701	Capturing Problem		
					CT2885	Physiological monitoring/measuring system transmitters/receivers/transceivers			A0701	Capturing Problem		
13892	2023	All other reportable incidents	36241	Pacemaker/defibrillator lead	CT1684	Cardiac pacemakers and associated devices	E2311	Discomfort	A070909	Oversensing		
									A072201	High Impedance		
							F08	Hospitalization or Prolonged Hospitalization	A070909	Oversensing		

									A072201	High Impedance		
							F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A070909	Oversensing		
									A072201	High Impedance		
							F1905	Device Revision or Replacement	A070909	Oversensing		
									A072201	High Impedance		
							F2203	Imaging Required	A070909	Oversensing		
									A072201	High Impedance		
13893	2023	All other reportable incidents	35223	Endocardial/interventricular septal pacing lead	CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070909	Oversensing		
							F26	No Health Consequences or Impact	A070909	Oversensing		
13894	2023	All other reportable incidents	37265	Dual-chamber implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E0623	Unspecified Heart Problem	A070909	Oversensing		
									A072201	High Impedance		
									A090801	Signal Artifact/Noise		
									A1301	Failure to Read Input Signal		
							F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A070909	Oversensing		
									A072201	High Impedance		
									A090801	Signal Artifact/Noise		
									A1301	Failure to Read Input Signal		
13895	2023	All other reportable incidents	47205	Cardiac pulse generator programmer	CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A150105	Key or Button Unresponsive/not Working		
							F26	No Health Consequences or Impact	A150105	Key or Button Unresponsive/not Working		
13896	2023	All other reportable incidents	47265	Dual-chamber implantable pacemaker, rate-responsive	CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	19/05/2023	19/05/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	19/05/2023	19/05/2023
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	19/05/2023	19/05/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	19/05/2023	19/05/2023
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	19/05/2023	19/05/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	19/05/2023	19/05/2023
13897	2023	All other reportable incidents	35223	Endocardial/interventricular septal pacing lead	CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A0401	Break		
									A070909	Oversensing		
							F23	Unexpected Medical Intervention	A0401	Break		
									A070909	Oversensing		
13898	2023	All other reportable incidents	47263	Cardiac resynchronization therapy implantable pacemaker	CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
13899	2023		47265	Dual-chamber implantable pacemaker, rate-responsive	CT141	Electrodes/Leads and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	17/02/2023	17/02/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	17/02/2023	17/02/2023
					CT145	Cardiovascular devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	17/02/2023	17/02/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	17/02/2023	17/02/2023
					CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	17/02/2023	17/02/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	17/02/2023	17/02/2023
					CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	17/02/2023	17/02/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	17/02/2023	17/02/2023
					CT233	Surgical	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	17/02/2023	17/02/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	17/02/2023	17/02/2023
					CT2406	Implantable	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	17/02/2023	17/02/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	17/02/2023	17/02/2023
					CT2407	Active implantable	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	17/02/2023	17/02/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	17/02/2023	17/02/2023
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	17/02/2023	17/02/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	17/02/2023	17/02/2023
					CT2727	Active implant devices/ancillaries	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	17/02/2023	17/02/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	17/02/2023	17/02/2023
					CT2750	Implantable cardiac electrical stimulators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	17/02/2023	17/02/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	17/02/2023	17/02/2023
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	17/02/2023	17/02/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	17/02/2023	17/02/2023
					CT2985	Active	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	17/02/2023	17/02/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	17/02/2023	17/02/2023
					CT3001	Implantable metal	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	17/02/2023	17/02/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	17/02/2023	17/02/2023
					CT3002	Implantable synthetic polymer	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	17/02/2023	17/02/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	17/02/2023	17/02/2023
					CT3006	Implantable inorganic material	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	17/02/2023	17/02/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	17/02/2023	17/02/2023
					CT314	Electrically powered	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	17/02/2023	17/02/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	17/02/2023	17/02/2023
					CT321	Long-term surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	17/02/2023	17/02/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	17/02/2023	17/02/2023
					CT449	Generators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	17/02/2023	17/02/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	17/02/2023	17/02/2023
					CT981	Single use	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	17/02/2023	17/02/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	17/02/2023	17/02/2023
					CT983	Surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	17/02/2023	17/02/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	17/02/2023	17/02/2023
13900	2023	All other reportable incidents	47265	Dual-chamber implantable pacemaker, rate-responsive	CT141	Electrodes/Leads and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070909	Oversensing		
									A0712	Pacing Problem		
									A090801	Signal Artifact/Noise		
							F1001	Absence of Treatment	A070909	Oversensing		
									A0712	Pacing Problem		
									A090801	Signal Artifact/Noise		
							F2203	Imaging Required	A070909	Oversensing		
									A0712	Pacing Problem		
									A090801	Signal Artifact/Noise		
					CT145	Cardiovascular devices	E2403	No Clinical Signs, Symptoms or Conditions	A070909	Oversensing		
									A0712	Pacing Problem		

								A090801	Signal Artifact/Noise		
					F1001	Absence of Treatment		A070909	Oversensing		
								A0712	Pacing Problem		
								A090801	Signal Artifact/Noise		
					F2203	Imaging Required		A070909	Oversensing		
								A0712	Pacing Problem		
								A090801	Signal Artifact/Noise		
				CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070909	Oversensing		
								A0712	Pacing Problem		
								A090801	Signal Artifact/Noise		
						F1001	Absence of Treatment	A070909	Oversensing		
								A0712	Pacing Problem		
								A090801	Signal Artifact/Noise		
						F2203	Imaging Required	A070909	Oversensing		
								A0712	Pacing Problem		
								A090801	Signal Artifact/Noise		
				CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A070909	Oversensing		
								A0712	Pacing Problem		
								A090801	Signal Artifact/Noise		
						F1001	Absence of Treatment	A070909	Oversensing		
								A0712	Pacing Problem		
								A090801	Signal Artifact/Noise		
						F2203	Imaging Required	A070909	Oversensing		
								A0712	Pacing Problem		
								A090801	Signal Artifact/Noise		
				CT233	Surgical	E2403	No Clinical Signs, Symptoms or Conditions	A070909	Oversensing		
								A0712	Pacing Problem		
								A090801	Signal Artifact/Noise		
						F1001	Absence of Treatment	A070909	Oversensing		
								A0712	Pacing Problem		
								A090801	Signal Artifact/Noise		
						F2203	Imaging Required	A070909	Oversensing		
								A0712	Pacing Problem		
								A090801	Signal Artifact/Noise		
				CT2406	Implantable	E2403	No Clinical Signs, Symptoms or Conditions	A070909	Oversensing		
								A0712	Pacing Problem		
								A090801	Signal Artifact/Noise		
						F1001	Absence of Treatment	A070909	Oversensing		
								A0712	Pacing Problem		
								A090801	Signal Artifact/Noise		
						F2203	Imaging Required	A070909	Oversensing		
								A0712	Pacing Problem		
								A090801	Signal Artifact/Noise		
				CT2407	Active implantable	E2403	No Clinical Signs, Symptoms or Conditions	A070909	Oversensing		
								A0712	Pacing Problem		
								A090801	Signal Artifact/Noise		
						F1001	Absence of Treatment	A070909	Oversensing		
								A0712	Pacing Problem		
								A090801	Signal Artifact/Noise		
						F2203	Imaging Required	A070909	Oversensing		
								A0712	Pacing Problem		
								A090801	Signal Artifact/Noise		
				CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A070909	Oversensing		
								A0712	Pacing Problem		
								A090801	Signal Artifact/Noise		
						F1001	Absence of Treatment	A070909	Oversensing		
								A0712	Pacing Problem		
								A090801	Signal Artifact/Noise		
						F2203	Imaging Required	A070909	Oversensing		
								A0712	Pacing Problem		
								A090801	Signal Artifact/Noise		
				CT2727	Active implant devices/ancillaries	E2403	No Clinical Signs, Symptoms or Conditions	A070909	Oversensing		
								A0712	Pacing Problem		
								A090801	Signal Artifact/Noise		
						F1001	Absence of Treatment	A070909	Oversensing		
								A0712	Pacing Problem		
								A090801	Signal Artifact/Noise		
						F2203	Imaging Required	A070909	Oversensing		
								A0712	Pacing Problem		
								A090801	Signal Artifact/Noise		
				CT2750	Implantable cardiac electrical stimulators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070909	Oversensing		
								A0712	Pacing Problem		
								A090801	Signal Artifact/Noise		
						F1001	Absence of Treatment	A070909	Oversensing		
								A0712	Pacing Problem		
								A090801	Signal Artifact/Noise		
						F2203	Imaging Required	A070909	Oversensing		
								A0712	Pacing Problem		
								A090801	Signal Artifact/Noise		
				CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A070909	Oversensing		
								A0712	Pacing Problem		
								A090801	Signal Artifact/Noise		
						F1001	Absence of Treatment	A070909	Oversensing		
								A0712	Pacing Problem		
								A090801	Signal Artifact/Noise		
						F2203	Imaging Required	A070909	Oversensing		
								A0712	Pacing Problem		
								A090801	Signal Artifact/Noise		

					CT2985	Active	E2403	No Clinical Signs, Symptoms or Conditions	A070909	Oversensing		
									A0712	Pacing Problem		
									A090801	Signal Artifact/Noise		
							F1001	Absence of Treatment	A070909	Oversensing		
									A0712	Pacing Problem		
									A090801	Signal Artifact/Noise		
							F2203	Imaging Required	A070909	Oversensing		
									A0712	Pacing Problem		
									A090801	Signal Artifact/Noise		
					CT3001	Implantable metal	E2403	No Clinical Signs, Symptoms or Conditions	A070909	Oversensing		
									A0712	Pacing Problem		
									A090801	Signal Artifact/Noise		
							F1001	Absence of Treatment	A070909	Oversensing		
									A0712	Pacing Problem		
									A090801	Signal Artifact/Noise		
							F2203	Imaging Required	A070909	Oversensing		
									A0712	Pacing Problem		
									A090801	Signal Artifact/Noise		
					CT3002	Implantable synthetic polymer	E2403	No Clinical Signs, Symptoms or Conditions	A070909	Oversensing		
									A0712	Pacing Problem		
									A090801	Signal Artifact/Noise		
							F1001	Absence of Treatment	A070909	Oversensing		
									A0712	Pacing Problem		
									A090801	Signal Artifact/Noise		
							F2203	Imaging Required	A070909	Oversensing		
									A0712	Pacing Problem		
									A090801	Signal Artifact/Noise		
					CT3006	Implantable inorganic material	E2403	No Clinical Signs, Symptoms or Conditions	A070909	Oversensing		
									A0712	Pacing Problem		
									A090801	Signal Artifact/Noise		
							F1001	Absence of Treatment	A070909	Oversensing		
									A0712	Pacing Problem		
									A090801	Signal Artifact/Noise		
							F2203	Imaging Required	A070909	Oversensing		
									A0712	Pacing Problem		
									A090801	Signal Artifact/Noise		
					CT314	Electrically powered	E2403	No Clinical Signs, Symptoms or Conditions	A070909	Oversensing		
									A0712	Pacing Problem		
									A090801	Signal Artifact/Noise		
							F1001	Absence of Treatment	A070909	Oversensing		
									A0712	Pacing Problem		
									A090801	Signal Artifact/Noise		
							F2203	Imaging Required	A070909	Oversensing		
									A0712	Pacing Problem		
									A090801	Signal Artifact/Noise		
					CT321	Long-term surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A070909	Oversensing		
									A0712	Pacing Problem		
									A090801	Signal Artifact/Noise		
							F1001	Absence of Treatment	A070909	Oversensing		
									A0712	Pacing Problem		
									A090801	Signal Artifact/Noise		
							F2203	Imaging Required	A070909	Oversensing		
									A0712	Pacing Problem		
									A090801	Signal Artifact/Noise		
					CT449	Generators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070909	Oversensing		
									A0712	Pacing Problem		
									A090801	Signal Artifact/Noise		
							F1001	Absence of Treatment	A070909	Oversensing		
									A0712	Pacing Problem		
									A090801	Signal Artifact/Noise		
							F2203	Imaging Required	A070909	Oversensing		
									A0712	Pacing Problem		
									A090801	Signal Artifact/Noise		
					CT981	Single use	E2403	No Clinical Signs, Symptoms or Conditions	A070909	Oversensing		
									A0712	Pacing Problem		
									A090801	Signal Artifact/Noise		
							F1001	Absence of Treatment	A070909	Oversensing		
									A0712	Pacing Problem		
									A090801	Signal Artifact/Noise		
							F2203	Imaging Required	A070909	Oversensing		
									A0712	Pacing Problem		
									A090801	Signal Artifact/Noise		
					CT983	Surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A070909	Oversensing		
									A0712	Pacing Problem		
									A090801	Signal Artifact/Noise		
							F1001	Absence of Treatment	A070909	Oversensing		
									A0712	Pacing Problem		
									A090801	Signal Artifact/Noise		
							F2203	Imaging Required	A070909	Oversensing		
									A0712	Pacing Problem		
									A090801	Signal Artifact/Noise		
13901	2023	All other reportable incidents	36241	Pacemaker/defibrillator lead	CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070909	Oversensing		
									A072201	High Impedance		
									A1301	Failure to Read Input Signal		
							F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A070909	Oversensing		
									A072201	High Impedance		
									A1301	Failure to Read Input Signal		
13902	2023	All other reportable incidents	36241	Pacemaker/defibrillator lead	CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	05/05/2023	05/05/2023

							F08	Hospitalization or Prolonged Hospitalization	A24	No Apparent Device Problem	05/05/2023	05/05/2023
							F1903	Device Explantation	A24	No Apparent Device Problem	05/05/2023	05/05/2023
13903	2023	All other reportable incidents	35852	Single-chamber implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	31/10/2023	31/10/2023
									A0908	Incorrect, Inadequate or Imprecise Result or Readings	31/10/2023	31/10/2023
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	31/10/2023	31/10/2023
									A0908	Incorrect, Inadequate or Imprecise Result or Readings	31/10/2023	31/10/2023
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	31/10/2023	31/10/2023
									A0908	Incorrect, Inadequate or Imprecise Result or Readings	31/10/2023	31/10/2023
13904	2023		47265	Dual-chamber implantable pacemaker, rate-responsive	CT141	Electrodes/Leads and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	12/07/2023	12/07/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	12/07/2023	12/07/2023
					CT145	Cardiovascular devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	12/07/2023	12/07/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	12/07/2023	12/07/2023
					CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	12/07/2023	12/07/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	12/07/2023	12/07/2023
					CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	12/07/2023	12/07/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	12/07/2023	12/07/2023
					CT233	Surgical	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	12/07/2023	12/07/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	12/07/2023	12/07/2023
					CT2406	Implantable	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	12/07/2023	12/07/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	12/07/2023	12/07/2023
					CT2407	Active implantable	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	12/07/2023	12/07/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	12/07/2023	12/07/2023
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	12/07/2023	12/07/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	12/07/2023	12/07/2023
					CT2727	Active implant devices/ancillaries	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	12/07/2023	12/07/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	12/07/2023	12/07/2023
					CT2750	Implantable cardiac electrical stimulators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	12/07/2023	12/07/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	12/07/2023	12/07/2023
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	12/07/2023	12/07/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	12/07/2023	12/07/2023
					CT2985	Active	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	12/07/2023	12/07/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	12/07/2023	12/07/2023
					CT3001	Implantable metal	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	12/07/2023	12/07/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	12/07/2023	12/07/2023
					CT3002	Implantable synthetic polymer	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	12/07/2023	12/07/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	12/07/2023	12/07/2023
					CT3006	Implantable inorganic material	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	12/07/2023	12/07/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	12/07/2023	12/07/2023
					CT314	Electrically powered	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	12/07/2023	12/07/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	12/07/2023	12/07/2023
					CT321	Long-term surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	12/07/2023	12/07/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	12/07/2023	12/07/2023
					CT449	Generators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	12/07/2023	12/07/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	12/07/2023	12/07/2023
					CT981	Single use	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	12/07/2023	12/07/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	12/07/2023	12/07/2023
					CT983	Surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	12/07/2023	12/07/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	12/07/2023	12/07/2023
13905	2023	All other reportable incidents	35223	Endocardial/interventricular septal pacing lead	CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070102	High Capture Threshold	26/05/2023	26/05/2023
									A072201	High Impedance	26/05/2023	26/05/2023
							F08	Hospitalization or Prolonged Hospitalization	A070102	High Capture Threshold	26/05/2023	26/05/2023
									A072201	High Impedance	26/05/2023	26/05/2023
							F12	Serious Injury/ Illness/ Impairment	A070102	High Capture Threshold	26/05/2023	26/05/2023
									A072201	High Impedance	26/05/2023	26/05/2023
							F1905	Device Revision or Replacement	A070102	High Capture Threshold	26/05/2023	26/05/2023
									A072201	High Impedance	26/05/2023	26/05/2023
							F23	Unexpected Medical Intervention	A070102	High Capture Threshold	26/05/2023	26/05/2023
									A072201	High Impedance	26/05/2023	26/05/2023
13906	2023	All other reportable incidents	36241	Pacemaker/defibrillator lead	CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A072201	High Impedance		
							F26	No Health Consequences or Impact	A072201	High Impedance		
13907	2023	All other reportable incidents	35223	Endocardial/interventricular septal pacing lead	CT1684	Cardiac pacemakers and associated devices	E0604	Cardiac Perforation	A24	No Apparent Device Problem		
							F12	Serious Injury/ Illness/ Impairment	A24	No Apparent Device Problem		
13908	2023	All other reportable incidents	47267	Single-chamber implantable pacemaker, rate-responsive	CT141	Electrodes/Leads and associated devices	E0112	Dizziness	A070909	Oversensing		
							F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A070909	Oversensing		
							F11	Non-Serious Injury/ Illness/ Impairment	A070909	Oversensing		
							F15	Recognized Device or Procedural Complication	A070909	Oversensing		
					CT145	Cardiovascular devices	E0112	Dizziness	A070909	Oversensing		
							F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A070909	Oversensing		
							F11	Non-Serious Injury/ Illness/ Impairment	A070909	Oversensing		
							F15	Recognized Device or Procedural Complication	A070909	Oversensing		
					CT1684	Cardiac pacemakers and associated devices	E0112	Dizziness	A070909	Oversensing		
							F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A070909	Oversensing		
							F11	Non-Serious Injury/ Illness/ Impairment	A070909	Oversensing		
							F15	Recognized Device or Procedural Complication	A070909	Oversensing		
					CT1685	Cardiac pacemakers	E0112	Dizziness	A070909	Oversensing		
							F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A070909	Oversensing		
							F11	Non-Serious Injury/ Illness/ Impairment	A070909	Oversensing		
							F15	Recognized Device or Procedural Complication	A070909	Oversensing		
					CT233	Surgical	E0112	Dizziness	A070909	Oversensing		
							F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A070909	Oversensing		
							F11	Non-Serious Injury/ Illness/ Impairment	A070909	Oversensing		
							F15	Recognized Device or Procedural Complication	A070909	Oversensing		
					CT2406	Implantable	E0112	Dizziness	A070909	Oversensing		
							F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A070909	Oversensing		
							F11	Non-Serious Injury/ Illness/ Impairment	A070909	Oversensing		

							F15	Recognized Device or Procedural Complication	A070909	Oversensing		
					CT2407	Active implantable	E0112	Dizziness	A070909	Oversensing		
							F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A070909	Oversensing		
							F11	Non-Serious Injury/ Illness/ Impairment	A070909	Oversensing		
							F15	Recognized Device or Procedural Complication	A070909	Oversensing		
					CT257	Implantable cardiac pacemakers	E0112	Dizziness	A070909	Oversensing		
							F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A070909	Oversensing		
							F11	Non-Serious Injury/ Illness/ Impairment	A070909	Oversensing		
							F15	Recognized Device or Procedural Complication	A070909	Oversensing		
					CT2727	Active implant devices/ancillaries	E0112	Dizziness	A070909	Oversensing		
							F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A070909	Oversensing		
							F11	Non-Serious Injury/ Illness/ Impairment	A070909	Oversensing		
							F15	Recognized Device or Procedural Complication	A070909	Oversensing		
					CT2750	Implantable cardiac electrical stimulators and associated devices	E0112	Dizziness	A070909	Oversensing		
							F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A070909	Oversensing		
							F11	Non-Serious Injury/ Illness/ Impairment	A070909	Oversensing		
							F15	Recognized Device or Procedural Complication	A070909	Oversensing		
					CT2866	Implantable cardiac electrical stimulators	E0112	Dizziness	A070909	Oversensing		
							F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A070909	Oversensing		
							F11	Non-Serious Injury/ Illness/ Impairment	A070909	Oversensing		
							F15	Recognized Device or Procedural Complication	A070909	Oversensing		
					CT2985	Active	E0112	Dizziness	A070909	Oversensing		
							F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A070909	Oversensing		
							F11	Non-Serious Injury/ Illness/ Impairment	A070909	Oversensing		
							F15	Recognized Device or Procedural Complication	A070909	Oversensing		
					CT3001	Implantable metal	E0112	Dizziness	A070909	Oversensing		
							F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A070909	Oversensing		
							F11	Non-Serious Injury/ Illness/ Impairment	A070909	Oversensing		
							F15	Recognized Device or Procedural Complication	A070909	Oversensing		
					CT3002	Implantable synthetic polymer	E0112	Dizziness	A070909	Oversensing		
							F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A070909	Oversensing		
							F11	Non-Serious Injury/ Illness/ Impairment	A070909	Oversensing		
							F15	Recognized Device or Procedural Complication	A070909	Oversensing		
					CT3006	Implantable inorganic material	E0112	Dizziness	A070909	Oversensing		
							F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A070909	Oversensing		
							F11	Non-Serious Injury/ Illness/ Impairment	A070909	Oversensing		
							F15	Recognized Device or Procedural Complication	A070909	Oversensing		
					CT314	Electrically powered	E0112	Dizziness	A070909	Oversensing		
							F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A070909	Oversensing		
							F11	Non-Serious Injury/ Illness/ Impairment	A070909	Oversensing		
							F15	Recognized Device or Procedural Complication	A070909	Oversensing		
					CT321	Long-term surgical invasive	E0112	Dizziness	A070909	Oversensing		
							F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A070909	Oversensing		
							F11	Non-Serious Injury/ Illness/ Impairment	A070909	Oversensing		
							F15	Recognized Device or Procedural Complication	A070909	Oversensing		
					CT449	Generators and associated devices	E0112	Dizziness	A070909	Oversensing		
							F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A070909	Oversensing		
							F11	Non-Serious Injury/ Illness/ Impairment	A070909	Oversensing		
							F15	Recognized Device or Procedural Complication	A070909	Oversensing		
					CT981	Single use	E0112	Dizziness	A070909	Oversensing		
							F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A070909	Oversensing		
							F11	Non-Serious Injury/ Illness/ Impairment	A070909	Oversensing		
							F15	Recognized Device or Procedural Complication	A070909	Oversensing		
					CT983	Surgical invasive	E0112	Dizziness	A070909	Oversensing		
							F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A070909	Oversensing		
							F11	Non-Serious Injury/ Illness/ Impairment	A070909	Oversensing		
							F15	Recognized Device or Procedural Complication	A070909	Oversensing		
13909	2023	All other reportable incidents	36241	Pacemaker/defibrillator lead	CT1684	Cardiac pacemakers and associated devices	E0112	Dizziness	A070909	Oversensing		
							F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A070909	Oversensing		
							F11	Non-Serious Injury/ Illness/ Impairment	A070909	Oversensing		
							F15	Recognized Device or Procedural Complication	A070909	Oversensing		
13910	2023	All other reportable incidents	47265	Dual-chamber implantable pacemaker, rate-responsive	CT141	Electrodes/Leads and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	02/06/2023	02/06/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	02/06/2023	02/06/2023
					CT145	Cardiovascular devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	02/06/2023	02/06/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	02/06/2023	02/06/2023
					CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	02/06/2023	02/06/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	02/06/2023	02/06/2023
					CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	02/06/2023	02/06/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	02/06/2023	02/06/2023
					CT233	Surgical	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	02/06/2023	02/06/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	02/06/2023	02/06/2023
					CT2406	Implantable	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	02/06/2023	02/06/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	02/06/2023	02/06/2023
					CT2407	Active implantable	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	02/06/2023	02/06/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	02/06/2023	02/06/2023
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	02/06/2023	02/06/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	02/06/2023	02/06/2023
					CT2727	Active implant devices/ancillaries	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	02/06/2023	02/06/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	02/06/2023	02/06/2023
					CT2750	Implantable cardiac electrical stimulators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	02/06/2023	02/06/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	02/06/2023	02/06/2023
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	02/06/2023	02/06/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	02/06/2023	02/06/2023
					CT2985	Active	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	02/06/2023	02/06/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	02/06/2023	02/06/2023
					CT3001	Implantable metal	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	02/06/2023	02/06/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	02/06/2023	02/06/2023
					CT3002	Implantable synthetic polymer	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	02/06/2023	02/06/2023

							F1901	Additional Surgery	A24	No Apparent Device Problem	02/06/2023	02/06/2023
					CT3006	Implantable inorganic material	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	02/06/2023	02/06/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	02/06/2023	02/06/2023
					CT314	Electrically powered	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	02/06/2023	02/06/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	02/06/2023	02/06/2023
					CT321	Long-term surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	02/06/2023	02/06/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	02/06/2023	02/06/2023
					CT449	Generators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	02/06/2023	02/06/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	02/06/2023	02/06/2023
					CT981	Single use	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	02/06/2023	02/06/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	02/06/2023	02/06/2023
					CT983	Surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	02/06/2023	02/06/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	02/06/2023	02/06/2023
13911	2023		47265	Dual-chamber implantable pacemaker, rate-responsive	CT141	Electrodes/Leads and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	10/12/2022	10/12/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	10/12/2022	10/12/2022
					CT145	Cardiovascular devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	10/12/2022	10/12/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	10/12/2022	10/12/2022
					CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	10/12/2022	10/12/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	10/12/2022	10/12/2022
					CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	10/12/2022	10/12/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	10/12/2022	10/12/2022
					CT233	Surgical	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	10/12/2022	10/12/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	10/12/2022	10/12/2022
					CT2406	Implantable	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	10/12/2022	10/12/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	10/12/2022	10/12/2022
					CT2407	Active implantable	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	10/12/2022	10/12/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	10/12/2022	10/12/2022
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	10/12/2022	10/12/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	10/12/2022	10/12/2022
					CT2727	Active implant devices/ancillaries	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	10/12/2022	10/12/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	10/12/2022	10/12/2022
					CT2750	Implantable cardiac electrical stimulators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	10/12/2022	10/12/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	10/12/2022	10/12/2022
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	10/12/2022	10/12/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	10/12/2022	10/12/2022
					CT2985	Active	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	10/12/2022	10/12/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	10/12/2022	10/12/2022
					CT3001	Implantable metal	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	10/12/2022	10/12/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	10/12/2022	10/12/2022
					CT3002	Implantable synthetic polymer	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	10/12/2022	10/12/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	10/12/2022	10/12/2022
					CT3006	Implantable inorganic material	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	10/12/2022	10/12/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	10/12/2022	10/12/2022
					CT314	Electrically powered	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	10/12/2022	10/12/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	10/12/2022	10/12/2022
					CT321	Long-term surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	10/12/2022	10/12/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	10/12/2022	10/12/2022
					CT449	Generators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	10/12/2022	10/12/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	10/12/2022	10/12/2022
					CT981	Single use	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	10/12/2022	10/12/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	10/12/2022	10/12/2022
					CT983	Surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	10/12/2022	10/12/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	10/12/2022	10/12/2022
13912	2023	All other reportable incidents	17882	Physiologic-monitoring defibrillation system	CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A1603	Failure of Device to Self-Test		
							F27	Problem Identified During Non-Clinical Procedure	A1603	Failure of Device to Self-Test		
13913	2023	All other reportable incidents	47265	Dual-chamber implantable pacemaker, rate-responsive	CT141	Electrodes/Leads and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	16/08/2023	16/08/2023
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	16/08/2023	16/08/2023
							F12	Serious Injury/ Illness/ Impairment	A070504	Premature Discharge of Battery	16/08/2023	16/08/2023
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	16/08/2023	16/08/2023
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery	16/08/2023	16/08/2023
					CT145	Cardiovascular devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	16/08/2023	16/08/2023
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	16/08/2023	16/08/2023
							F12	Serious Injury/ Illness/ Impairment	A070504	Premature Discharge of Battery	16/08/2023	16/08/2023
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	16/08/2023	16/08/2023
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery	16/08/2023	16/08/2023
					CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	16/08/2023	16/08/2023
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	16/08/2023	16/08/2023
							F12	Serious Injury/ Illness/ Impairment	A070504	Premature Discharge of Battery	16/08/2023	16/08/2023
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	16/08/2023	16/08/2023
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery	16/08/2023	16/08/2023
					CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	16/08/2023	16/08/2023
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	16/08/2023	16/08/2023
							F12	Serious Injury/ Illness/ Impairment	A070504	Premature Discharge of Battery	16/08/2023	16/08/2023
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	16/08/2023	16/08/2023
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery	16/08/2023	16/08/2023
					CT233	Surgical	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	16/08/2023	16/08/2023
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	16/08/2023	16/08/2023
							F12	Serious Injury/ Illness/ Impairment	A070504	Premature Discharge of Battery	16/08/2023	16/08/2023
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	16/08/2023	16/08/2023
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery	16/08/2023	16/08/2023
					CT2406	Implantable	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	16/08/2023	16/08/2023
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	16/08/2023	16/08/2023
							F12	Serious Injury/ Illness/ Impairment	A070504	Premature Discharge of Battery	16/08/2023	16/08/2023
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	16/08/2023	16/08/2023
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery	16/08/2023	16/08/2023
					CT2407	Active implantable	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	16/08/2023	16/08/2023
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	16/08/2023	16/08/2023
							F12	Serious Injury/ Illness/ Impairment	A070504	Premature Discharge of Battery	16/08/2023	16/08/2023

							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	16/08/2023	16/08/2023
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery	16/08/2023	16/08/2023
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	16/08/2023	16/08/2023
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	16/08/2023	16/08/2023
							F12	Serious Injury/ Illness/ Impairment	A070504	Premature Discharge of Battery	16/08/2023	16/08/2023
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	16/08/2023	16/08/2023
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery	16/08/2023	16/08/2023
					CT2727	Active implant devices/ancillaries	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	16/08/2023	16/08/2023
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	16/08/2023	16/08/2023
							F12	Serious Injury/ Illness/ Impairment	A070504	Premature Discharge of Battery	16/08/2023	16/08/2023
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	16/08/2023	16/08/2023
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery	16/08/2023	16/08/2023
					CT2750	Implantable cardiac electrical stimulators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	16/08/2023	16/08/2023
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	16/08/2023	16/08/2023
							F12	Serious Injury/ Illness/ Impairment	A070504	Premature Discharge of Battery	16/08/2023	16/08/2023
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	16/08/2023	16/08/2023
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery	16/08/2023	16/08/2023
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	16/08/2023	16/08/2023
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	16/08/2023	16/08/2023
							F12	Serious Injury/ Illness/ Impairment	A070504	Premature Discharge of Battery	16/08/2023	16/08/2023
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	16/08/2023	16/08/2023
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery	16/08/2023	16/08/2023
					CT2985	Active	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	16/08/2023	16/08/2023
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	16/08/2023	16/08/2023
							F12	Serious Injury/ Illness/ Impairment	A070504	Premature Discharge of Battery	16/08/2023	16/08/2023
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	16/08/2023	16/08/2023
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery	16/08/2023	16/08/2023
					CT3001	Implantable metal	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	16/08/2023	16/08/2023
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	16/08/2023	16/08/2023
							F12	Serious Injury/ Illness/ Impairment	A070504	Premature Discharge of Battery	16/08/2023	16/08/2023
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	16/08/2023	16/08/2023
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery	16/08/2023	16/08/2023
					CT3002	Implantable synthetic polymer	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	16/08/2023	16/08/2023
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	16/08/2023	16/08/2023
							F12	Serious Injury/ Illness/ Impairment	A070504	Premature Discharge of Battery	16/08/2023	16/08/2023
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	16/08/2023	16/08/2023
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery	16/08/2023	16/08/2023
					CT3006	Implantable inorganic material	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	16/08/2023	16/08/2023
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	16/08/2023	16/08/2023
							F12	Serious Injury/ Illness/ Impairment	A070504	Premature Discharge of Battery	16/08/2023	16/08/2023
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	16/08/2023	16/08/2023
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery	16/08/2023	16/08/2023
					CT314	Electrically powered	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	16/08/2023	16/08/2023
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	16/08/2023	16/08/2023
							F12	Serious Injury/ Illness/ Impairment	A070504	Premature Discharge of Battery	16/08/2023	16/08/2023
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	16/08/2023	16/08/2023
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery	16/08/2023	16/08/2023
					CT321	Long-term surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	16/08/2023	16/08/2023
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	16/08/2023	16/08/2023
							F12	Serious Injury/ Illness/ Impairment	A070504	Premature Discharge of Battery	16/08/2023	16/08/2023
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	16/08/2023	16/08/2023
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery	16/08/2023	16/08/2023
					CT449	Generators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	16/08/2023	16/08/2023
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	16/08/2023	16/08/2023
							F12	Serious Injury/ Illness/ Impairment	A070504	Premature Discharge of Battery	16/08/2023	16/08/2023
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	16/08/2023	16/08/2023
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery	16/08/2023	16/08/2023
					CT981	Single use	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	16/08/2023	16/08/2023
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	16/08/2023	16/08/2023
							F12	Serious Injury/ Illness/ Impairment	A070504	Premature Discharge of Battery	16/08/2023	16/08/2023
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	16/08/2023	16/08/2023
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery	16/08/2023	16/08/2023
					CT983	Surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	16/08/2023	16/08/2023
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	16/08/2023	16/08/2023
							F12	Serious Injury/ Illness/ Impairment	A070504	Premature Discharge of Battery	16/08/2023	16/08/2023
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	16/08/2023	16/08/2023
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery	16/08/2023	16/08/2023
13914	2023	Death	47267	Single-chamber implantable pacemaker, rate-responsive	CT1685	Cardiac pacemakers	E2401	Insufficient Information	A0705	Battery Problem		
							F02	Death	A0705	Battery Problem		
					CT257	Implantable cardiac pacemakers	E2401	Insufficient Information	A0705	Battery Problem		
							F02	Death	A0705	Battery Problem		
					CT2866	Implantable cardiac electrical stimulators	E2401	Insufficient Information	A0705	Battery Problem		
							F02	Death	A0705	Battery Problem		
13915	2023	All other reportable incidents	47265	Dual-chamber implantable pacemaker, rate-responsive	CT1685	Cardiac pacemakers	E0119	Loss of Consciousness	A070101	Failure to Capture	30/10/2023	30/10/2023
									A090404	Output Above Specifications	30/10/2023	30/10/2023
							E011902	Presyncope	A070101	Failure to Capture	30/10/2023	30/10/2023
									A090404	Output Above Specifications	30/10/2023	30/10/2023
							E0601	Arrhythmia	A070101	Failure to Capture	30/10/2023	30/10/2023
									A090404	Output Above Specifications	30/10/2023	30/10/2023
							E060104	Bradycardia	A070101	Failure to Capture	30/10/2023	30/10/2023
									A090404	Output Above Specifications	30/10/2023	30/10/2023
							F08	Hospitalization or Prolonged Hospitalization	A070101	Failure to Capture	30/10/2023	30/10/2023
									A090404	Output Above Specifications	30/10/2023	30/10/2023
							F1203	Life Threatening Illness or Injury	A070101	Failure to Capture	30/10/2023	30/10/2023
									A090404	Output Above Specifications	30/10/2023	30/10/2023
							F1905	Device Revision or Replacement	A070101	Failure to Capture	30/10/2023	30/10/2023
									A090404	Output Above Specifications	30/10/2023	30/10/2023
							F23	Unexpected Medical Intervention	A070101	Failure to Capture	30/10/2023	30/10/2023

									A090404	Output Above Specifications	30/10/2023	30/10/2023
					CT257	Implantable cardiac pacemakers	E0119	Loss of Consciousness	A070101	Failure to Capture	30/10/2023	30/10/2023
									A090404	Output Above Specifications	30/10/2023	30/10/2023
							E011902	Presyncope	A070101	Failure to Capture	30/10/2023	30/10/2023
									A090404	Output Above Specifications	30/10/2023	30/10/2023
							E0601	Arrhythmia	A070101	Failure to Capture	30/10/2023	30/10/2023
									A090404	Output Above Specifications	30/10/2023	30/10/2023
							E060104	Bradycardia	A070101	Failure to Capture	30/10/2023	30/10/2023
									A090404	Output Above Specifications	30/10/2023	30/10/2023
							F08	Hospitalization or Prolonged Hospitalization	A070101	Failure to Capture	30/10/2023	30/10/2023
									A090404	Output Above Specifications	30/10/2023	30/10/2023
							F1203	Life Threatening Illness or Injury	A070101	Failure to Capture	30/10/2023	30/10/2023
									A090404	Output Above Specifications	30/10/2023	30/10/2023
							F1905	Device Revision or Replacement	A070101	Failure to Capture	30/10/2023	30/10/2023
									A090404	Output Above Specifications	30/10/2023	30/10/2023
							F23	Unexpected Medical Intervention	A070101	Failure to Capture	30/10/2023	30/10/2023
									A090404	Output Above Specifications	30/10/2023	30/10/2023
					CT2866	Implantable cardiac electrical stimulators	E0119	Loss of Consciousness	A070101	Failure to Capture	30/10/2023	30/10/2023
									A090404	Output Above Specifications	30/10/2023	30/10/2023
							E011902	Presyncope	A070101	Failure to Capture	30/10/2023	30/10/2023
									A090404	Output Above Specifications	30/10/2023	30/10/2023
							E0601	Arrhythmia	A070101	Failure to Capture	30/10/2023	30/10/2023
									A090404	Output Above Specifications	30/10/2023	30/10/2023
							E060104	Bradycardia	A070101	Failure to Capture	30/10/2023	30/10/2023
									A090404	Output Above Specifications	30/10/2023	30/10/2023
							F08	Hospitalization or Prolonged Hospitalization	A070101	Failure to Capture	30/10/2023	30/10/2023
									A090404	Output Above Specifications	30/10/2023	30/10/2023
							F1203	Life Threatening Illness or Injury	A070101	Failure to Capture	30/10/2023	30/10/2023
									A090404	Output Above Specifications	30/10/2023	30/10/2023
							F1905	Device Revision or Replacement	A070101	Failure to Capture	30/10/2023	30/10/2023
									A090404	Output Above Specifications	30/10/2023	30/10/2023
							F23	Unexpected Medical Intervention	A070101	Failure to Capture	30/10/2023	30/10/2023
									A090404	Output Above Specifications	30/10/2023	30/10/2023
13916	2023	All other reportable incidents	47265	Dual-chamber implantable pacemaker, rate-responsive	CT141	Electrodes/Leads and associated devices	E0601	Arrhythmia	A0712	Pacing Problem		
									A071202	Inaccurate Synchronization		
							E0623	Unspecified Heart Problem	A0712	Pacing Problem		
									A071202	Inaccurate Synchronization		
							F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A0712	Pacing Problem		
									A071202	Inaccurate Synchronization		
					CT145	Cardiovascular devices	E0601	Arrhythmia	A0712	Pacing Problem		
									A071202	Inaccurate Synchronization		
							E0623	Unspecified Heart Problem	A0712	Pacing Problem		
									A071202	Inaccurate Synchronization		
							F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A0712	Pacing Problem		
									A071202	Inaccurate Synchronization		
					CT1684	Cardiac pacemakers and associated devices	E0601	Arrhythmia	A0712	Pacing Problem		
									A071202	Inaccurate Synchronization		
							E0623	Unspecified Heart Problem	A0712	Pacing Problem		
									A071202	Inaccurate Synchronization		
							F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A0712	Pacing Problem		
									A071202	Inaccurate Synchronization		
					CT1685	Cardiac pacemakers	E0601	Arrhythmia	A0712	Pacing Problem		
									A071202	Inaccurate Synchronization		
							E0623	Unspecified Heart Problem	A0712	Pacing Problem		
									A071202	Inaccurate Synchronization		
							F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A0712	Pacing Problem		
									A071202	Inaccurate Synchronization		
					CT233	Surgical	E0601	Arrhythmia	A0712	Pacing Problem		
									A071202	Inaccurate Synchronization		
							E0623	Unspecified Heart Problem	A0712	Pacing Problem		
									A071202	Inaccurate Synchronization		
							F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A0712	Pacing Problem		
									A071202	Inaccurate Synchronization		
					CT2406	Implantable	E0601	Arrhythmia	A0712	Pacing Problem		
									A071202	Inaccurate Synchronization		
							E0623	Unspecified Heart Problem	A0712	Pacing Problem		
									A071202	Inaccurate Synchronization		
							F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A0712	Pacing Problem		
									A071202	Inaccurate Synchronization		
					CT2407	Active implantable	E0601	Arrhythmia	A0712	Pacing Problem		
									A071202	Inaccurate Synchronization		
							E0623	Unspecified Heart Problem	A0712	Pacing Problem		
									A071202	Inaccurate Synchronization		
							F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A0712	Pacing Problem		
									A071202	Inaccurate Synchronization		
					CT257	Implantable cardiac pacemakers	E0601	Arrhythmia	A0712	Pacing Problem		
									A071202	Inaccurate Synchronization		
							E0623	Unspecified Heart Problem	A0712	Pacing Problem		
									A071202	Inaccurate Synchronization		
							F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A0712	Pacing Problem		
									A071202	Inaccurate Synchronization		
					CT2727	Active implant devices/ancillaries	E0601	Arrhythmia	A0712	Pacing Problem		
									A071202	Inaccurate Synchronization		
							E0623	Unspecified Heart Problem	A0712	Pacing Problem		
									A071202	Inaccurate Synchronization		
							F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A0712	Pacing Problem		
									A071202	Inaccurate Synchronization		
					CT2750	Implantable cardiac electrical stimulators and associated devices	E0601	Arrhythmia	A0712	Pacing Problem		

									A071202	Inaccurate Synchronization		
							E0623	Unspecified Heart Problem	A0712	Pacing Problem		
									A071202	Inaccurate Synchronization		
							F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A0712	Pacing Problem		
									A071202	Inaccurate Synchronization		
					CT2866	Implantable cardiac electrical stimulators	E0601	Arrhythmia	A0712	Pacing Problem		
									A071202	Inaccurate Synchronization		
							E0623	Unspecified Heart Problem	A0712	Pacing Problem		
									A071202	Inaccurate Synchronization		
							F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A0712	Pacing Problem		
									A071202	Inaccurate Synchronization		
					CT2985	Active	E0601	Arrhythmia	A0712	Pacing Problem		
									A071202	Inaccurate Synchronization		
							E0623	Unspecified Heart Problem	A0712	Pacing Problem		
									A071202	Inaccurate Synchronization		
							F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A0712	Pacing Problem		
									A071202	Inaccurate Synchronization		
					CT3001	Implantable metal	E0601	Arrhythmia	A0712	Pacing Problem		
									A071202	Inaccurate Synchronization		
							E0623	Unspecified Heart Problem	A0712	Pacing Problem		
									A071202	Inaccurate Synchronization		
							F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A0712	Pacing Problem		
									A071202	Inaccurate Synchronization		
					CT3002	Implantable synthetic polymer	E0601	Arrhythmia	A0712	Pacing Problem		
									A071202	Inaccurate Synchronization		
							E0623	Unspecified Heart Problem	A0712	Pacing Problem		
									A071202	Inaccurate Synchronization		
							F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A0712	Pacing Problem		
									A071202	Inaccurate Synchronization		
					CT3006	Implantable inorganic material	E0601	Arrhythmia	A0712	Pacing Problem		
									A071202	Inaccurate Synchronization		
							E0623	Unspecified Heart Problem	A0712	Pacing Problem		
									A071202	Inaccurate Synchronization		
							F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A0712	Pacing Problem		
									A071202	Inaccurate Synchronization		
					CT314	Electrically powered	E0601	Arrhythmia	A0712	Pacing Problem		
									A071202	Inaccurate Synchronization		
							E0623	Unspecified Heart Problem	A0712	Pacing Problem		
									A071202	Inaccurate Synchronization		
							F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A0712	Pacing Problem		
									A071202	Inaccurate Synchronization		
					CT321	Long-term surgical invasive	E0601	Arrhythmia	A0712	Pacing Problem		
									A071202	Inaccurate Synchronization		
							E0623	Unspecified Heart Problem	A0712	Pacing Problem		
									A071202	Inaccurate Synchronization		
							F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A0712	Pacing Problem		
									A071202	Inaccurate Synchronization		
					CT449	Generators and associated devices	E0601	Arrhythmia	A0712	Pacing Problem		
									A071202	Inaccurate Synchronization		
							E0623	Unspecified Heart Problem	A0712	Pacing Problem		
									A071202	Inaccurate Synchronization		
							F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A0712	Pacing Problem		
									A071202	Inaccurate Synchronization		
					CT981	Single use	E0601	Arrhythmia	A0712	Pacing Problem		
									A071202	Inaccurate Synchronization		
							E0623	Unspecified Heart Problem	A0712	Pacing Problem		
									A071202	Inaccurate Synchronization		
							F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A0712	Pacing Problem		
									A071202	Inaccurate Synchronization		
					CT983	Surgical invasive	E0601	Arrhythmia	A0712	Pacing Problem		
									A071202	Inaccurate Synchronization		
							E0623	Unspecified Heart Problem	A0712	Pacing Problem		
									A071202	Inaccurate Synchronization		
							F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A0712	Pacing Problem		
									A071202	Inaccurate Synchronization		
13917	2023	All other reportable incidents	35852	Single-chamber implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery		
							F12	Serious Injury/ Illness/ Impairment	A070504	Premature Discharge of Battery		
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
13918	2023	All other reportable incidents	35852	Single-chamber implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A0712	Pacing Problem		
									A1301	Failure to Read Input Signal		
							F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A0712	Pacing Problem		
									A1301	Failure to Read Input Signal		
13919	2023	All other reportable incidents	35852	Single-chamber implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E2104	Electric Shock	A070909	Oversensing	01/12/2022	01/12/2022
									A090801	Signal Artifact/Noise	01/12/2022	01/12/2022
									A1301	Failure to Read Input Signal	01/12/2022	01/12/2022
							F08	Hospitalization or Prolonged Hospitalization	A070909	Oversensing	01/12/2022	01/12/2022
									A090801	Signal Artifact/Noise	01/12/2022	01/12/2022
									A1301	Failure to Read Input Signal	01/12/2022	01/12/2022
							F1905	Device Revision or Replacement	A070909	Oversensing	01/12/2022	01/12/2022
									A090801	Signal Artifact/Noise	01/12/2022	01/12/2022
									A1301	Failure to Read Input Signal	01/12/2022	01/12/2022
							F2203	Imaging Required	A070909	Oversensing	01/12/2022	01/12/2022
									A090801	Signal Artifact/Noise	01/12/2022	01/12/2022
									A1301	Failure to Read Input Signal	01/12/2022	01/12/2022
13920	2023	All other reportable incidents	35223	Endocardial/interventricular septal pacing lead	CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070909	Oversensing		
									A0722	Impedance Problem		

							F12	Serious Injury/ Illness/ Impairment	A070909	Oversensing		
									A0722	Impedance Problem		
							F23	Unexpected Medical Intervention	A070909	Oversensing		
									A0722	Impedance Problem		
13921	2023	All other reportable incidents	35852	Single-chamber implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
13922	2023	All other reportable incidents	35223	Endocardial/interventricular septal pacing lead	CT1684	Cardiac pacemakers and associated devices	E2311	Discomfort	A070101	Failure to Capture		
									A070909	Oversensing		
									A072202	Low Impedance		
									A090407	Therapy Delivered to Incorrect Body Area		
							F1901	Additional Surgery	A070101	Failure to Capture		
									A070909	Oversensing		
									A072202	Low Impedance		
									A090407	Therapy Delivered to Incorrect Body Area		
13923	2023	All other reportable incidents	47270	Cardiac resynchronization therapy implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E060109	Tachycardia	A0701	Capturing Problem		
							F08	Hospitalization or Prolonged Hospitalization	A0701	Capturing Problem		
							F12	Serious Injury/ Illness/ Impairment	A0701	Capturing Problem		
							F23	Unexpected Medical Intervention	A0701	Capturing Problem		
13924	2023	All other reportable incidents	17882	Physiologic-monitoring defibrillation system	CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem		
							F27	Problem Identified During Non-Clinical Procedure	A0705	Battery Problem		
13925	2023		17882	Physiologic-monitoring defibrillation system	CT1684	Cardiac pacemakers and associated devices	E06	Heart	A07	Electrical /Electronic Property Problem		
									A0713	Defibrillation/Stimulation Problem		
									A27	Appropriate Device Problem Term/Code not Available		
							F26	No Health Consequences or Impact	A07	Electrical /Electronic Property Problem		
									A0713	Defibrillation/Stimulation Problem		
									A27	Appropriate Device Problem Term/Code not Available		
13926	2023		35223	Endocardial/interventricular septal pacing lead	CT1684	Cardiac pacemakers and associated devices	F08	Hospitalization or Prolonged Hospitalization	A07	Electrical /Electronic Property Problem		
13927	2023	All other reportable incidents	47265	Dual-chamber implantable pacemaker, rate-responsive	CT141	Electrodes/Leads and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A0203	Defective Device		
							F26	No Health Consequences or Impact	A0203	Defective Device		
					CT145	Cardiovascular devices	E2403	No Clinical Signs, Symptoms or Conditions	A0203	Defective Device		
							F26	No Health Consequences or Impact	A0203	Defective Device		
					CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A0203	Defective Device		
							F26	No Health Consequences or Impact	A0203	Defective Device		
					CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A0203	Defective Device		
							F26	No Health Consequences or Impact	A0203	Defective Device		
					CT233	Surgical	E2403	No Clinical Signs, Symptoms or Conditions	A0203	Defective Device		
							F26	No Health Consequences or Impact	A0203	Defective Device		
					CT2406	Implantable	E2403	No Clinical Signs, Symptoms or Conditions	A0203	Defective Device		
							F26	No Health Consequences or Impact	A0203	Defective Device		
					CT2407	Active implantable	E2403	No Clinical Signs, Symptoms or Conditions	A0203	Defective Device		
							F26	No Health Consequences or Impact	A0203	Defective Device		
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A0203	Defective Device		
							F26	No Health Consequences or Impact	A0203	Defective Device		
					CT2727	Active implant devices/ancillaries	E2403	No Clinical Signs, Symptoms or Conditions	A0203	Defective Device		
							F26	No Health Consequences or Impact	A0203	Defective Device		
					CT2750	Implantable cardiac electrical stimulators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A0203	Defective Device		
							F26	No Health Consequences or Impact	A0203	Defective Device		
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A0203	Defective Device		
							F26	No Health Consequences or Impact	A0203	Defective Device		
					CT2985	Active	E2403	No Clinical Signs, Symptoms or Conditions	A0203	Defective Device		
							F26	No Health Consequences or Impact	A0203	Defective Device		
					CT3001	Implantable metal	E2403	No Clinical Signs, Symptoms or Conditions	A0203	Defective Device		
							F26	No Health Consequences or Impact	A0203	Defective Device		
					CT3002	Implantable synthetic polymer	E2403	No Clinical Signs, Symptoms or Conditions	A0203	Defective Device		
							F26	No Health Consequences or Impact	A0203	Defective Device		
					CT3006	Implantable inorganic material	E2403	No Clinical Signs, Symptoms or Conditions	A0203	Defective Device		
							F26	No Health Consequences or Impact	A0203	Defective Device		
					CT314	Electrically powered	E2403	No Clinical Signs, Symptoms or Conditions	A0203	Defective Device		
							F26	No Health Consequences or Impact	A0203	Defective Device		
					CT321	Long-term surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A0203	Defective Device		
							F26	No Health Consequences or Impact	A0203	Defective Device		
					CT449	Generators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A0203	Defective Device		
							F26	No Health Consequences or Impact	A0203	Defective Device		
					CT981	Single use	E2403	No Clinical Signs, Symptoms or Conditions	A0203	Defective Device		
							F26	No Health Consequences or Impact	A0203	Defective Device		
					CT983	Surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A0203	Defective Device		
							F26	No Health Consequences or Impact	A0203	Defective Device		
13928	2023	All other reportable incidents	47265	Dual-chamber implantable pacemaker, rate-responsive	CT141	Electrodes/Leads and associated devices	E0306	Sepsis	A070101	Failure to Capture		
									A0712	Pacing Problem		
									A1301	Failure to Read Input Signal		
							F08	Hospitalization or Prolonged Hospitalization	A070101	Failure to Capture		
									A0712	Pacing Problem		
									A1301	Failure to Read Input Signal		
							F2203	Imaging Required	A070101	Failure to Capture		
									A0712	Pacing Problem		
									A1301	Failure to Read Input Signal		
					CT145	Cardiovascular devices	E0306	Sepsis	A070101	Failure to Capture		
									A0712	Pacing Problem		
									A1301	Failure to Read Input Signal		
							F08	Hospitalization or Prolonged Hospitalization	A070101	Failure to Capture		
									A0712	Pacing Problem		
									A1301	Failure to Read Input Signal		
							F2203	Imaging Required	A070101	Failure to Capture		

									A0712	Pacing Problem		
									A1301	Failure to Read Input Signal		
					CT1684	Cardiac pacemakers and associated devices	E0306	Sepsis	A070101	Failure to Capture		
									A0712	Pacing Problem		
									A1301	Failure to Read Input Signal		
							F08	Hospitalization or Prolonged Hospitalization	A070101	Failure to Capture		
									A0712	Pacing Problem		
									A1301	Failure to Read Input Signal		
							F2203	Imaging Required	A070101	Failure to Capture		
									A0712	Pacing Problem		
									A1301	Failure to Read Input Signal		
					CT1685	Cardiac pacemakers	E0306	Sepsis	A070101	Failure to Capture		
									A0712	Pacing Problem		
									A1301	Failure to Read Input Signal		
							F08	Hospitalization or Prolonged Hospitalization	A070101	Failure to Capture		
									A0712	Pacing Problem		
									A1301	Failure to Read Input Signal		
							F2203	Imaging Required	A070101	Failure to Capture		
									A0712	Pacing Problem		
									A1301	Failure to Read Input Signal		
					CT233	Surgical	E0306	Sepsis	A070101	Failure to Capture		
									A0712	Pacing Problem		
									A1301	Failure to Read Input Signal		
							F08	Hospitalization or Prolonged Hospitalization	A070101	Failure to Capture		
									A0712	Pacing Problem		
									A1301	Failure to Read Input Signal		
							F2203	Imaging Required	A070101	Failure to Capture		
									A0712	Pacing Problem		
									A1301	Failure to Read Input Signal		
					CT2406	Implantable	E0306	Sepsis	A070101	Failure to Capture		
									A0712	Pacing Problem		
									A1301	Failure to Read Input Signal		
							F08	Hospitalization or Prolonged Hospitalization	A070101	Failure to Capture		
									A0712	Pacing Problem		
									A1301	Failure to Read Input Signal		
							F2203	Imaging Required	A070101	Failure to Capture		
									A0712	Pacing Problem		
									A1301	Failure to Read Input Signal		
					CT2407	Active implantable	E0306	Sepsis	A070101	Failure to Capture		
									A0712	Pacing Problem		
									A1301	Failure to Read Input Signal		
							F08	Hospitalization or Prolonged Hospitalization	A070101	Failure to Capture		
									A0712	Pacing Problem		
									A1301	Failure to Read Input Signal		
							F2203	Imaging Required	A070101	Failure to Capture		
									A0712	Pacing Problem		
									A1301	Failure to Read Input Signal		
					CT257	Implantable cardiac pacemakers	E0306	Sepsis	A070101	Failure to Capture		
									A0712	Pacing Problem		
									A1301	Failure to Read Input Signal		
							F08	Hospitalization or Prolonged Hospitalization	A070101	Failure to Capture		
									A0712	Pacing Problem		
									A1301	Failure to Read Input Signal		
							F2203	Imaging Required	A070101	Failure to Capture		
									A0712	Pacing Problem		
									A1301	Failure to Read Input Signal		
					CT2727	Active implant devices/ancillaries	E0306	Sepsis	A070101	Failure to Capture		
									A0712	Pacing Problem		
									A1301	Failure to Read Input Signal		
							F08	Hospitalization or Prolonged Hospitalization	A070101	Failure to Capture		
									A0712	Pacing Problem		
									A1301	Failure to Read Input Signal		
							F2203	Imaging Required	A070101	Failure to Capture		
									A0712	Pacing Problem		
									A1301	Failure to Read Input Signal		
					CT2750	Implantable cardiac electrical stimulators and associated devices	E0306	Sepsis	A070101	Failure to Capture		
									A0712	Pacing Problem		
									A1301	Failure to Read Input Signal		
							F08	Hospitalization or Prolonged Hospitalization	A070101	Failure to Capture		
									A0712	Pacing Problem		
									A1301	Failure to Read Input Signal		
							F2203	Imaging Required	A070101	Failure to Capture		
									A0712	Pacing Problem		
									A1301	Failure to Read Input Signal		
					CT2866	Implantable cardiac electrical stimulators	E0306	Sepsis	A070101	Failure to Capture		
									A0712	Pacing Problem		
									A1301	Failure to Read Input Signal		
							F08	Hospitalization or Prolonged Hospitalization	A070101	Failure to Capture		
									A0712	Pacing Problem		
									A1301	Failure to Read Input Signal		
							F2203	Imaging Required	A070101	Failure to Capture		
									A0712	Pacing Problem		
									A1301	Failure to Read Input Signal		
					CT2985	Active	E0306	Sepsis	A070101	Failure to Capture		
									A0712	Pacing Problem		
									A1301	Failure to Read Input Signal		
							F08	Hospitalization or Prolonged Hospitalization	A070101	Failure to Capture		
									A0712	Pacing Problem		

									A1301	Failure to Read Input Signal		
							F2203	Imaging Required	A070101	Failure to Capture		
									A0712	Pacing Problem		
									A1301	Failure to Read Input Signal		
					CT3001	Implantable metal	E0306	Sepsis	A070101	Failure to Capture		
									A0712	Pacing Problem		
									A1301	Failure to Read Input Signal		
							F08	Hospitalization or Prolonged Hospitalization	A070101	Failure to Capture		
									A0712	Pacing Problem		
									A1301	Failure to Read Input Signal		
							F2203	Imaging Required	A070101	Failure to Capture		
									A0712	Pacing Problem		
									A1301	Failure to Read Input Signal		
					CT3002	Implantable synthetic polymer	E0306	Sepsis	A070101	Failure to Capture		
									A0712	Pacing Problem		
									A1301	Failure to Read Input Signal		
							F08	Hospitalization or Prolonged Hospitalization	A070101	Failure to Capture		
									A0712	Pacing Problem		
									A1301	Failure to Read Input Signal		
							F2203	Imaging Required	A070101	Failure to Capture		
									A0712	Pacing Problem		
									A1301	Failure to Read Input Signal		
					CT3006	Implantable inorganic material	E0306	Sepsis	A070101	Failure to Capture		
									A0712	Pacing Problem		
									A1301	Failure to Read Input Signal		
							F08	Hospitalization or Prolonged Hospitalization	A070101	Failure to Capture		
									A0712	Pacing Problem		
									A1301	Failure to Read Input Signal		
							F2203	Imaging Required	A070101	Failure to Capture		
									A0712	Pacing Problem		
									A1301	Failure to Read Input Signal		
					CT314	Electrically powered	E0306	Sepsis	A070101	Failure to Capture		
									A0712	Pacing Problem		
									A1301	Failure to Read Input Signal		
							F08	Hospitalization or Prolonged Hospitalization	A070101	Failure to Capture		
									A0712	Pacing Problem		
									A1301	Failure to Read Input Signal		
							F2203	Imaging Required	A070101	Failure to Capture		
									A0712	Pacing Problem		
									A1301	Failure to Read Input Signal		
					CT321	Long-term surgical invasive	E0306	Sepsis	A070101	Failure to Capture		
									A0712	Pacing Problem		
									A1301	Failure to Read Input Signal		
							F08	Hospitalization or Prolonged Hospitalization	A070101	Failure to Capture		
									A0712	Pacing Problem		
									A1301	Failure to Read Input Signal		
							F2203	Imaging Required	A070101	Failure to Capture		
									A0712	Pacing Problem		
									A1301	Failure to Read Input Signal		
					CT449	Generators and associated devices	E0306	Sepsis	A070101	Failure to Capture		
									A0712	Pacing Problem		
									A1301	Failure to Read Input Signal		
							F08	Hospitalization or Prolonged Hospitalization	A070101	Failure to Capture		
									A0712	Pacing Problem		
									A1301	Failure to Read Input Signal		
							F2203	Imaging Required	A070101	Failure to Capture		
									A0712	Pacing Problem		
									A1301	Failure to Read Input Signal		
					CT981	Single use	E0306	Sepsis	A070101	Failure to Capture		
									A0712	Pacing Problem		
									A1301	Failure to Read Input Signal		
							F08	Hospitalization or Prolonged Hospitalization	A070101	Failure to Capture		
									A0712	Pacing Problem		
									A1301	Failure to Read Input Signal		
							F2203	Imaging Required	A070101	Failure to Capture		
									A0712	Pacing Problem		
									A1301	Failure to Read Input Signal		
					CT983	Surgical invasive	E0306	Sepsis	A070101	Failure to Capture		
									A0712	Pacing Problem		
									A1301	Failure to Read Input Signal		
							F08	Hospitalization or Prolonged Hospitalization	A070101	Failure to Capture		
									A0712	Pacing Problem		
									A1301	Failure to Read Input Signal		
							F2203	Imaging Required	A070101	Failure to Capture		
									A0712	Pacing Problem		
									A1301	Failure to Read Input Signal		
13929	2023	All other reportable incidents	36241	Pacemaker/defibrillator lead	CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070909	Oversensing		
									A090801	Signal Artifact/Noise		
							F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A070909	Oversensing		
									A090801	Signal Artifact/Noise		
13930	2023	All other reportable incidents	35852	Single-chamber implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A0722	Impedance Problem		
									A072201	High Impedance		
							F26	No Health Consequences or Impact	A0722	Impedance Problem		
									A072201	High Impedance		
13931	2023	All other reportable incidents	36241	Pacemaker/defibrillator lead	CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A072201	High Impedance		
							F26	No Health Consequences or Impact	A072201	High Impedance		
13932	2023	All other reportable incidents	47267	Single-chamber implantable pacemaker, rate-responsive	CT141	Electrodes/Leads and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		

					CT145	Cardiovascular devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT233	Surgical	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT2406	Implantable	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT2407	Active implantable	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT2727	Active implant devices/ancillaries	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT2750	Implantable cardiac electrical stimulators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT2985	Active	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT3001	Implantable metal	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT3002	Implantable synthetic polymer	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT3006	Implantable inorganic material	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT314	Electrically powered	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT321	Long-term surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT449	Generators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT981	Single use	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT983	Surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
13933	2023	All other reportable incidents	47267	Single-chamber implantable pacemaker, rate-responsive	CT141	Electrodes/Leads and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
					CT145	Cardiovascular devices	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
					CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
					CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
					CT233	Surgical	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
					CT2406	Implantable	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
					CT2407	Active implantable	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
					CT2727	Active implant devices/ancillaries	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
					CT2750	Implantable cardiac electrical stimulators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
					CT2985	Active	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
					CT3001	Implantable metal	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		

							F26	No Health Consequences or Impact	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
					CT3002	Implantable synthetic polymer	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
					CT3006	Implantable inorganic material	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
					CT314	Electrically powered	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
					CT321	Long-term surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
					CT449	Generators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
					CT981	Single use	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
					CT983	Surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
13934	2023	All other reportable incidents	47265	Dual-chamber implantable pacemaker, rate-responsive	CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F24	Insufficient Health Impact Information	A070504	Premature Discharge of Battery		
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F24	Insufficient Health Impact Information	A070504	Premature Discharge of Battery		
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F24	Insufficient Health Impact Information	A070504	Premature Discharge of Battery		
13935	2023	All other reportable incidents	35223	Endocardial/interventricular septal pacing lead	CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A040101	Fracture		
									A070909	Oversensing		
							F1901	Additional Surgery	A040101	Fracture		
									A070909	Oversensing		
13936	2023	All other reportable incidents	47263	Cardiac resynchronization therapy implantable pacemaker	CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A0908	Incorrect, Inadequate or Imprecise Result or Readings		
							F26	No Health Consequences or Impact	A0908	Incorrect, Inadequate or Imprecise Result or Readings		
					CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A0908	Incorrect, Inadequate or Imprecise Result or Readings		
							F26	No Health Consequences or Impact	A0908	Incorrect, Inadequate or Imprecise Result or Readings		
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A0908	Incorrect, Inadequate or Imprecise Result or Readings		
							F26	No Health Consequences or Impact	A0908	Incorrect, Inadequate or Imprecise Result or Readings		
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A0908	Incorrect, Inadequate or Imprecise Result or Readings		
							F26	No Health Consequences or Impact	A0908	Incorrect, Inadequate or Imprecise Result or Readings		
13937	2023	All other reportable incidents	47263	Cardiac resynchronization therapy implantable pacemaker	CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070909	Oversensing		
									A0712	Pacing Problem		
									A090801	Signal Artifact/Noise		
									A1301	Failure to Read Input Signal		
							F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A070909	Oversensing		
									A0712	Pacing Problem		
									A090801	Signal Artifact/Noise		
									A1301	Failure to Read Input Signal		
					CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A070909	Oversensing		
									A0712	Pacing Problem		
									A090801	Signal Artifact/Noise		
									A1301	Failure to Read Input Signal		
							F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A070909	Oversensing		
									A0712	Pacing Problem		
									A090801	Signal Artifact/Noise		
									A1301	Failure to Read Input Signal		
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A070909	Oversensing		
									A0712	Pacing Problem		
									A090801	Signal Artifact/Noise		
									A1301	Failure to Read Input Signal		
							F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A070909	Oversensing		
									A0712	Pacing Problem		
									A090801	Signal Artifact/Noise		
									A1301	Failure to Read Input Signal		
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A070909	Oversensing		
									A0712	Pacing Problem		
									A090801	Signal Artifact/Noise		
									A1301	Failure to Read Input Signal		
							F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A070909	Oversensing		
									A0712	Pacing Problem		
									A090801	Signal Artifact/Noise		
									A1301	Failure to Read Input Signal		
13938	2023	All other reportable incidents	47270	Cardiac resynchronization therapy implantable	CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		

				defibrillator					A0908	Incorrect, Inadequate or Imprecise Result or Readings		
							F24	Insufficient Health Impact Information	A070504	Premature Discharge of Battery		
									A0908	Incorrect, Inadequate or Imprecise Result or Readings		
13939	2023	All other reportable incidents	47270	Cardiac resynchronization therapy implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A0203	Defective Device		
							F26	No Health Consequences or Impact	A0203	Defective Device		
13940	2023	All other reportable incidents	47267	Single-chamber implantable pacemaker, rate-responsive	CT141	Electrodes/Leads and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem	19/04/2023	19/04/2023
									A070504	Premature Discharge of Battery	19/04/2023	19/04/2023
							F08	Hospitalization or Prolonged Hospitalization	A0705	Battery Problem	19/04/2023	19/04/2023
									A070504	Premature Discharge of Battery	19/04/2023	19/04/2023
							F1905	Device Revision or Replacement	A0705	Battery Problem	19/04/2023	19/04/2023
									A070504	Premature Discharge of Battery	19/04/2023	19/04/2023
					CT145	Cardiovascular devices	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem	19/04/2023	19/04/2023
									A070504	Premature Discharge of Battery	19/04/2023	19/04/2023
							F08	Hospitalization or Prolonged Hospitalization	A0705	Battery Problem	19/04/2023	19/04/2023
									A070504	Premature Discharge of Battery	19/04/2023	19/04/2023
							F1905	Device Revision or Replacement	A0705	Battery Problem	19/04/2023	19/04/2023
									A070504	Premature Discharge of Battery	19/04/2023	19/04/2023
					CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem	19/04/2023	19/04/2023
									A070504	Premature Discharge of Battery	19/04/2023	19/04/2023
							F08	Hospitalization or Prolonged Hospitalization	A0705	Battery Problem	19/04/2023	19/04/2023
									A070504	Premature Discharge of Battery	19/04/2023	19/04/2023
							F1905	Device Revision or Replacement	A0705	Battery Problem	19/04/2023	19/04/2023
									A070504	Premature Discharge of Battery	19/04/2023	19/04/2023
					CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem	19/04/2023	19/04/2023
									A070504	Premature Discharge of Battery	19/04/2023	19/04/2023
							F08	Hospitalization or Prolonged Hospitalization	A0705	Battery Problem	19/04/2023	19/04/2023
									A070504	Premature Discharge of Battery	19/04/2023	19/04/2023
							F1905	Device Revision or Replacement	A0705	Battery Problem	19/04/2023	19/04/2023
									A070504	Premature Discharge of Battery	19/04/2023	19/04/2023
					CT233	Surgical	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem	19/04/2023	19/04/2023
									A070504	Premature Discharge of Battery	19/04/2023	19/04/2023
							F08	Hospitalization or Prolonged Hospitalization	A0705	Battery Problem	19/04/2023	19/04/2023
									A070504	Premature Discharge of Battery	19/04/2023	19/04/2023
							F1905	Device Revision or Replacement	A0705	Battery Problem	19/04/2023	19/04/2023
									A070504	Premature Discharge of Battery	19/04/2023	19/04/2023
					CT2406	Implantable	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem	19/04/2023	19/04/2023
									A070504	Premature Discharge of Battery	19/04/2023	19/04/2023
							F08	Hospitalization or Prolonged Hospitalization	A0705	Battery Problem	19/04/2023	19/04/2023
									A070504	Premature Discharge of Battery	19/04/2023	19/04/2023
							F1905	Device Revision or Replacement	A0705	Battery Problem	19/04/2023	19/04/2023
									A070504	Premature Discharge of Battery	19/04/2023	19/04/2023
					CT2407	Active implantable	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem	19/04/2023	19/04/2023
									A070504	Premature Discharge of Battery	19/04/2023	19/04/2023
							F08	Hospitalization or Prolonged Hospitalization	A0705	Battery Problem	19/04/2023	19/04/2023
									A070504	Premature Discharge of Battery	19/04/2023	19/04/2023
							F1905	Device Revision or Replacement	A0705	Battery Problem	19/04/2023	19/04/2023
									A070504	Premature Discharge of Battery	19/04/2023	19/04/2023
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem	19/04/2023	19/04/2023
									A070504	Premature Discharge of Battery	19/04/2023	19/04/2023
							F08	Hospitalization or Prolonged Hospitalization	A0705	Battery Problem	19/04/2023	19/04/2023
									A070504	Premature Discharge of Battery	19/04/2023	19/04/2023
							F1905	Device Revision or Replacement	A0705	Battery Problem	19/04/2023	19/04/2023
									A070504	Premature Discharge of Battery	19/04/2023	19/04/2023
					CT2727	Active implant devices/ancillaries	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem	19/04/2023	19/04/2023
									A070504	Premature Discharge of Battery	19/04/2023	19/04/2023
							F08	Hospitalization or Prolonged Hospitalization	A0705	Battery Problem	19/04/2023	19/04/2023
									A070504	Premature Discharge of Battery	19/04/2023	19/04/2023
							F1905	Device Revision or Replacement	A0705	Battery Problem	19/04/2023	19/04/2023
									A070504	Premature Discharge of Battery	19/04/2023	19/04/2023
					CT2750	Implantable cardiac electrical stimulators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem	19/04/2023	19/04/2023
									A070504	Premature Discharge of Battery	19/04/2023	19/04/2023
							F08	Hospitalization or Prolonged Hospitalization	A0705	Battery Problem	19/04/2023	19/04/2023
									A070504	Premature Discharge of Battery	19/04/2023	19/04/2023
							F1905	Device Revision or Replacement	A0705	Battery Problem	19/04/2023	19/04/2023
									A070504	Premature Discharge of Battery	19/04/2023	19/04/2023
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem	19/04/2023	19/04/2023
									A070504	Premature Discharge of Battery	19/04/2023	19/04/2023
							F08	Hospitalization or Prolonged Hospitalization	A0705	Battery Problem	19/04/2023	19/04/2023
									A070504	Premature Discharge of Battery	19/04/2023	19/04/2023
							F1905	Device Revision or Replacement	A0705	Battery Problem	19/04/2023	19/04/2023
									A070504	Premature Discharge of Battery	19/04/2023	19/04/2023
					CT2985	Active	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem	19/04/2023	19/04/2023
									A070504	Premature Discharge of Battery	19/04/2023	19/04/2023
							F08	Hospitalization or Prolonged Hospitalization	A0705	Battery Problem	19/04/2023	19/04/2023
									A070504	Premature Discharge of Battery	19/04/2023	19/04/2023
							F1905	Device Revision or Replacement	A0705	Battery Problem	19/04/2023	19/04/2023
									A070504	Premature Discharge of Battery	19/04/2023	19/04/2023
					CT3001	Implantable metal	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem	19/04/2023	19/04/2023
									A070504	Premature Discharge of Battery	19/04/2023	19/04/2023
							F08	Hospitalization or Prolonged Hospitalization	A0705	Battery Problem	19/04/2023	19/04/2023
									A070504	Premature Discharge of Battery	19/04/2023	19/04/2023
							F1905	Device Revision or Replacement	A0705	Battery Problem	19/04/2023	19/04/2023
									A070504	Premature Discharge of Battery	19/04/2023	19/04/2023
					CT3002	Implantable synthetic polymer	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem	19/04/2023	19/04/2023
									A070504	Premature Discharge of Battery	19/04/2023	19/04/2023
							F08	Hospitalization or Prolonged Hospitalization	A0705	Battery Problem	19/04/2023	19/04/2023

								A070504	Premature Discharge of Battery	19/04/2023	19/04/2023
						F1905	Device Revision or Replacement	A0705	Battery Problem	19/04/2023	19/04/2023
								A070504	Premature Discharge of Battery	19/04/2023	19/04/2023
					CT3006	Implantable inorganic material	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem	19/04/2023	19/04/2023
									A070504	Premature Discharge of Battery	19/04/2023	19/04/2023
							F08	Hospitalization or Prolonged Hospitalization	A0705	Battery Problem	19/04/2023	19/04/2023
									A070504	Premature Discharge of Battery	19/04/2023	19/04/2023
							F1905	Device Revision or Replacement	A0705	Battery Problem	19/04/2023	19/04/2023
									A070504	Premature Discharge of Battery	19/04/2023	19/04/2023
					CT314	Electrically powered	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem	19/04/2023	19/04/2023
									A070504	Premature Discharge of Battery	19/04/2023	19/04/2023
							F08	Hospitalization or Prolonged Hospitalization	A0705	Battery Problem	19/04/2023	19/04/2023
									A070504	Premature Discharge of Battery	19/04/2023	19/04/2023
							F1905	Device Revision or Replacement	A0705	Battery Problem	19/04/2023	19/04/2023
									A070504	Premature Discharge of Battery	19/04/2023	19/04/2023
					CT321	Long-term surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem	19/04/2023	19/04/2023
									A070504	Premature Discharge of Battery	19/04/2023	19/04/2023
							F08	Hospitalization or Prolonged Hospitalization	A0705	Battery Problem	19/04/2023	19/04/2023
									A070504	Premature Discharge of Battery	19/04/2023	19/04/2023
							F1905	Device Revision or Replacement	A0705	Battery Problem	19/04/2023	19/04/2023
									A070504	Premature Discharge of Battery	19/04/2023	19/04/2023
					CT449	Generators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem	19/04/2023	19/04/2023
									A070504	Premature Discharge of Battery	19/04/2023	19/04/2023
							F08	Hospitalization or Prolonged Hospitalization	A0705	Battery Problem	19/04/2023	19/04/2023
									A070504	Premature Discharge of Battery	19/04/2023	19/04/2023
							F1905	Device Revision or Replacement	A0705	Battery Problem	19/04/2023	19/04/2023
									A070504	Premature Discharge of Battery	19/04/2023	19/04/2023
					CT981	Single use	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem	19/04/2023	19/04/2023
									A070504	Premature Discharge of Battery	19/04/2023	19/04/2023
							F08	Hospitalization or Prolonged Hospitalization	A0705	Battery Problem	19/04/2023	19/04/2023
									A070504	Premature Discharge of Battery	19/04/2023	19/04/2023
							F1905	Device Revision or Replacement	A0705	Battery Problem	19/04/2023	19/04/2023
									A070504	Premature Discharge of Battery	19/04/2023	19/04/2023
					CT983	Surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem	19/04/2023	19/04/2023
									A070504	Premature Discharge of Battery	19/04/2023	19/04/2023
							F08	Hospitalization or Prolonged Hospitalization	A0705	Battery Problem	19/04/2023	19/04/2023
									A070504	Premature Discharge of Battery	19/04/2023	19/04/2023
							F1905	Device Revision or Replacement	A0705	Battery Problem	19/04/2023	19/04/2023
									A070504	Premature Discharge of Battery	19/04/2023	19/04/2023
13941	2023	All other reportable incidents	47265	Dual-chamber implantable pacemaker, rate-responsive	CT1685	Cardiac pacemakers	E2311	Discomfort	A071102	Failure to Interrogate	16/08/2023	16/08/2023
									A071204	No Pacing	16/08/2023	16/08/2023
							F1901	Additional Surgery	A071102	Failure to Interrogate	16/08/2023	16/08/2023
									A071204	No Pacing	16/08/2023	16/08/2023
					CT257	Implantable cardiac pacemakers	E2311	Discomfort	A071102	Failure to Interrogate	16/08/2023	16/08/2023
									A071204	No Pacing	16/08/2023	16/08/2023
							F1901	Additional Surgery	A071102	Failure to Interrogate	16/08/2023	16/08/2023
									A071204	No Pacing	16/08/2023	16/08/2023
					CT2866	Implantable cardiac electrical stimulators	E2311	Discomfort	A071102	Failure to Interrogate	16/08/2023	16/08/2023
									A071204	No Pacing	16/08/2023	16/08/2023
							F1901	Additional Surgery	A071102	Failure to Interrogate	16/08/2023	16/08/2023
									A071204	No Pacing	16/08/2023	16/08/2023
13942	2023	All other reportable incidents	36241	Pacemaker/defibrillator lead	CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070101	Failure to Capture		
									A070102	High Capture Threshold		
							F26	No Health Consequences or Impact	A070101	Failure to Capture		
									A070102	High Capture Threshold		
13943	2023		35852	Single-chamber implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E0601	Arrhythmia	A071301	Failure to Deliver Shock/Stimulation	24/08/2023	24/08/2023
							F1901	Additional Surgery	A071301	Failure to Deliver Shock/Stimulation	24/08/2023	24/08/2023
13944	2023	All other reportable incidents	47265	Dual-chamber implantable pacemaker, rate-responsive	CT141	Electrodes/Leads and associated devices	E060101	Asystole	A0709	Device Sensing Problem		
									A1301	Failure to Read Input Signal		
							F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A0709	Device Sensing Problem		
									A1301	Failure to Read Input Signal		
					CT145	Cardiovascular devices	E060101	Asystole	A0709	Device Sensing Problem		
									A1301	Failure to Read Input Signal		
							F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A0709	Device Sensing Problem		
									A1301	Failure to Read Input Signal		
					CT1684	Cardiac pacemakers and associated devices	E060101	Asystole	A0709	Device Sensing Problem		
									A1301	Failure to Read Input Signal		
							F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A0709	Device Sensing Problem		
									A1301	Failure to Read Input Signal		
					CT1685	Cardiac pacemakers	E060101	Asystole	A0709	Device Sensing Problem		
									A1301	Failure to Read Input Signal		
							F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A0709	Device Sensing Problem		
									A1301	Failure to Read Input Signal		
					CT233	Surgical	E060101	Asystole	A0709	Device Sensing Problem		
									A1301	Failure to Read Input Signal		
							F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A0709	Device Sensing Problem		
									A1301	Failure to Read Input Signal		
					CT2406	Implantable	E060101	Asystole	A0709	Device Sensing Problem		
									A1301	Failure to Read Input Signal		
							F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A0709	Device Sensing Problem		
									A1301	Failure to Read Input Signal		
					CT2407	Active implantable	E060101	Asystole	A0709	Device Sensing Problem		
									A1301	Failure to Read Input Signal		
							F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A0709	Device Sensing Problem		
									A1301	Failure to Read Input Signal		
					CT257	Implantable cardiac pacemakers	E060101	Asystole	A0709	Device Sensing Problem		
									A1301	Failure to Read Input Signal		
							F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A0709	Device Sensing Problem		

									A1301	Failure to Read Input Signal		
					CT2727	Active implant devices/ancillaries	E060101	Asystole	A0709	Device Sensing Problem		
									A1301	Failure to Read Input Signal		
							F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A0709	Device Sensing Problem		
									A1301	Failure to Read Input Signal		
					CT2750	Implantable cardiac electrical stimulators and associated devices	E060101	Asystole	A0709	Device Sensing Problem		
									A1301	Failure to Read Input Signal		
							F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A0709	Device Sensing Problem		
									A1301	Failure to Read Input Signal		
					CT2866	Implantable cardiac electrical stimulators	E060101	Asystole	A0709	Device Sensing Problem		
									A1301	Failure to Read Input Signal		
							F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A0709	Device Sensing Problem		
									A1301	Failure to Read Input Signal		
					CT2985	Active	E060101	Asystole	A0709	Device Sensing Problem		
									A1301	Failure to Read Input Signal		
							F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A0709	Device Sensing Problem		
									A1301	Failure to Read Input Signal		
					CT3001	Implantable metal	E060101	Asystole	A0709	Device Sensing Problem		
									A1301	Failure to Read Input Signal		
							F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A0709	Device Sensing Problem		
									A1301	Failure to Read Input Signal		
					CT3002	Implantable synthetic polymer	E060101	Asystole	A0709	Device Sensing Problem		
									A1301	Failure to Read Input Signal		
							F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A0709	Device Sensing Problem		
									A1301	Failure to Read Input Signal		
					CT3006	Implantable inorganic material	E060101	Asystole	A0709	Device Sensing Problem		
									A1301	Failure to Read Input Signal		
							F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A0709	Device Sensing Problem		
									A1301	Failure to Read Input Signal		
					CT314	Electrically powered	E060101	Asystole	A0709	Device Sensing Problem		
									A1301	Failure to Read Input Signal		
							F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A0709	Device Sensing Problem		
									A1301	Failure to Read Input Signal		
					CT321	Long-term surgical invasive	E060101	Asystole	A0709	Device Sensing Problem		
									A1301	Failure to Read Input Signal		
							F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A0709	Device Sensing Problem		
									A1301	Failure to Read Input Signal		
					CT449	Generators and associated devices	E060101	Asystole	A0709	Device Sensing Problem		
									A1301	Failure to Read Input Signal		
							F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A0709	Device Sensing Problem		
									A1301	Failure to Read Input Signal		
					CT981	Single use	E060101	Asystole	A0709	Device Sensing Problem		
									A1301	Failure to Read Input Signal		
							F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A0709	Device Sensing Problem		
									A1301	Failure to Read Input Signal		
					CT983	Surgical invasive	E060101	Asystole	A0709	Device Sensing Problem		
									A1301	Failure to Read Input Signal		
							F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A0709	Device Sensing Problem		
									A1301	Failure to Read Input Signal		
13945	2023		47265	Dual-chamber implantable pacemaker, rate-responsive	CT141	Electrodes/Leads and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	24/08/2023	24/08/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	24/08/2023	24/08/2023
					CT145	Cardiovascular devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	24/08/2023	24/08/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	24/08/2023	24/08/2023
					CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	24/08/2023	24/08/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	24/08/2023	24/08/2023
					CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	24/08/2023	24/08/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	24/08/2023	24/08/2023
					CT233	Surgical	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	24/08/2023	24/08/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	24/08/2023	24/08/2023
					CT2406	Implantable	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	24/08/2023	24/08/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	24/08/2023	24/08/2023
					CT2407	Active implantable	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	24/08/2023	24/08/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	24/08/2023	24/08/2023
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	24/08/2023	24/08/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	24/08/2023	24/08/2023
					CT2727	Active implant devices/ancillaries	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	24/08/2023	24/08/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	24/08/2023	24/08/2023
					CT2750	Implantable cardiac electrical stimulators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	24/08/2023	24/08/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	24/08/2023	24/08/2023
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	24/08/2023	24/08/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	24/08/2023	24/08/2023
					CT2985	Active	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	24/08/2023	24/08/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	24/08/2023	24/08/2023
					CT3001	Implantable metal	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	24/08/2023	24/08/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	24/08/2023	24/08/2023
					CT3002	Implantable synthetic polymer	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	24/08/2023	24/08/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	24/08/2023	24/08/2023
					CT3006	Implantable inorganic material	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	24/08/2023	24/08/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	24/08/2023	24/08/2023
					CT314	Electrically powered	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	24/08/2023	24/08/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	24/08/2023	24/08/2023
					CT321	Long-term surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	24/08/2023	24/08/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	24/08/2023	24/08/2023
					CT449	Generators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	24/08/2023	24/08/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	24/08/2023	24/08/2023
					CT981	Single use	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	24/08/2023	24/08/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	24/08/2023	24/08/2023
					CT983	Surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	24/08/2023	24/08/2023

							F1901	Additional Surgery	A24	No Apparent Device Problem	24/08/2023	24/08/2023
13946	2023	All other reportable incidents	36241	Pacemaker/defibrillator lead	CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070909	Oversensing		
									A090801	Signal Artifact/Noise		
									A1301	Failure to Read Input Signal		
							F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A070909	Oversensing		
									A090801	Signal Artifact/Noise		
									A1301	Failure to Read Input Signal		
13947	2023	All other reportable incidents	17882	Physiologic-monitoring defibrillation system	CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A1603	Failure of Device to Self-Test		
							F27	Problem Identified During Non-Clinical Procedure	A1603	Failure of Device to Self-Test		
13948	2023	All other reportable incidents	17882	Physiologic-monitoring defibrillation system	CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A0721	Circuit Failure		
							F27	Problem Identified During Non-Clinical Procedure	A0721	Circuit Failure		
13949	2023	All other reportable incidents	47267	Single-chamber implantable pacemaker, rate-responsive	CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem	29/06/2023	29/06/2023
									A070504	Premature Discharge of Battery	29/06/2023	29/06/2023
							F08	Hospitalization or Prolonged Hospitalization	A0705	Battery Problem	29/06/2023	29/06/2023
									A070504	Premature Discharge of Battery	29/06/2023	29/06/2023
							F1905	Device Revision or Replacement	A0705	Battery Problem	29/06/2023	29/06/2023
									A070504	Premature Discharge of Battery	29/06/2023	29/06/2023
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem	29/06/2023	29/06/2023
									A070504	Premature Discharge of Battery	29/06/2023	29/06/2023
							F08	Hospitalization or Prolonged Hospitalization	A0705	Battery Problem	29/06/2023	29/06/2023
									A070504	Premature Discharge of Battery	29/06/2023	29/06/2023
							F1905	Device Revision or Replacement	A0705	Battery Problem	29/06/2023	29/06/2023
									A070504	Premature Discharge of Battery	29/06/2023	29/06/2023
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem	29/06/2023	29/06/2023
									A070504	Premature Discharge of Battery	29/06/2023	29/06/2023
							F08	Hospitalization or Prolonged Hospitalization	A0705	Battery Problem	29/06/2023	29/06/2023
									A070504	Premature Discharge of Battery	29/06/2023	29/06/2023
							F1905	Device Revision or Replacement	A0705	Battery Problem	29/06/2023	29/06/2023
									A070504	Premature Discharge of Battery	29/06/2023	29/06/2023
13950	2023	All other reportable incidents	47265	Dual-chamber implantable pacemaker, rate-responsive	CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	18/08/2023	18/08/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	18/08/2023	18/08/2023
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	18/08/2023	18/08/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	18/08/2023	18/08/2023
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	18/08/2023	18/08/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	18/08/2023	18/08/2023
13951	2023	All other reportable incidents	17882	Physiologic-monitoring defibrillation system	CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070803	Failure to Power Up		
							F27	Problem Identified During Non-Clinical Procedure	A070803	Failure to Power Up		
13952	2023	All other reportable incidents	47265	Dual-chamber implantable pacemaker, rate-responsive	CT141	Electrodes/Leads and associated devices	E060109	Tachycardia	A070910	Undersensing		
									A0712	Pacing Problem		
									A071202	Inaccurate Synchronization		
							F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A070910	Undersensing		
									A0712	Pacing Problem		
									A071202	Inaccurate Synchronization		
							F15	Recognized Device or Procedural Complication	A070910	Undersensing		
									A0712	Pacing Problem		
									A071202	Inaccurate Synchronization		
					CT145	Cardiovascular devices	E060109	Tachycardia	A070910	Undersensing		
									A0712	Pacing Problem		
									A071202	Inaccurate Synchronization		
							F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A070910	Undersensing		
									A0712	Pacing Problem		
									A071202	Inaccurate Synchronization		
							F15	Recognized Device or Procedural Complication	A070910	Undersensing		
									A0712	Pacing Problem		
									A071202	Inaccurate Synchronization		
					CT1684	Cardiac pacemakers and associated devices	E060109	Tachycardia	A070910	Undersensing		
									A0712	Pacing Problem		
									A071202	Inaccurate Synchronization		
							F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A070910	Undersensing		
									A0712	Pacing Problem		
									A071202	Inaccurate Synchronization		
							F15	Recognized Device or Procedural Complication	A070910	Undersensing		
									A0712	Pacing Problem		
									A071202	Inaccurate Synchronization		
					CT1685	Cardiac pacemakers	E060109	Tachycardia	A070910	Undersensing		
									A0712	Pacing Problem		
									A071202	Inaccurate Synchronization		
							F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A070910	Undersensing		
									A0712	Pacing Problem		
									A071202	Inaccurate Synchronization		
							F15	Recognized Device or Procedural Complication	A070910	Undersensing		
									A0712	Pacing Problem		
									A071202	Inaccurate Synchronization		
					CT233	Surgical	E060109	Tachycardia	A070910	Undersensing		
									A0712	Pacing Problem		
									A071202	Inaccurate Synchronization		
							F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A070910	Undersensing		
									A0712	Pacing Problem		
									A071202	Inaccurate Synchronization		
							F15	Recognized Device or Procedural Complication	A070910	Undersensing		
									A0712	Pacing Problem		
									A071202	Inaccurate Synchronization		
					CT2406	Implantable	E060109	Tachycardia	A070910	Undersensing		
									A0712	Pacing Problem		
									A071202	Inaccurate Synchronization		
							F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A070910	Undersensing		
									A0712	Pacing Problem		
									A071202	Inaccurate Synchronization		

							F15	Recognized Device or Procedural Complication	A070910	Undersensing		
									A0712	Pacing Problem		
									A071202	Inaccurate Synchronization		
						CT2407	Active implantable	E060109	Tachycardia	A070910	Undersensing		
										A0712	Pacing Problem		
										A071202	Inaccurate Synchronization		
								F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A070910	Undersensing		
										A0712	Pacing Problem		
										A071202	Inaccurate Synchronization		
								F15	Recognized Device or Procedural Complication	A070910	Undersensing		
										A0712	Pacing Problem		
										A071202	Inaccurate Synchronization		
						CT257	Implantable cardiac pacemakers	E060109	Tachycardia	A070910	Undersensing		
										A0712	Pacing Problem		
										A071202	Inaccurate Synchronization		
								F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A070910	Undersensing		
										A0712	Pacing Problem		
										A071202	Inaccurate Synchronization		
								F15	Recognized Device or Procedural Complication	A070910	Undersensing		
										A0712	Pacing Problem		
										A071202	Inaccurate Synchronization		
						CT2727	Active implant devices/ancillaries	E060109	Tachycardia	A070910	Undersensing		
										A0712	Pacing Problem		
										A071202	Inaccurate Synchronization		
								F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A070910	Undersensing		
										A0712	Pacing Problem		
										A071202	Inaccurate Synchronization		
								F15	Recognized Device or Procedural Complication	A070910	Undersensing		
										A0712	Pacing Problem		
										A071202	Inaccurate Synchronization		
						CT2750	Implantable cardiac electrical stimulators and associated devices	E060109	Tachycardia	A070910	Undersensing		
										A0712	Pacing Problem		
										A071202	Inaccurate Synchronization		
								F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A070910	Undersensing		
										A0712	Pacing Problem		
										A071202	Inaccurate Synchronization		
								F15	Recognized Device or Procedural Complication	A070910	Undersensing		
										A0712	Pacing Problem		
										A071202	Inaccurate Synchronization		
						CT2866	Implantable cardiac electrical stimulators	E060109	Tachycardia	A070910	Undersensing		
										A0712	Pacing Problem		
										A071202	Inaccurate Synchronization		
								F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A070910	Undersensing		
										A0712	Pacing Problem		
										A071202	Inaccurate Synchronization		
								F15	Recognized Device or Procedural Complication	A070910	Undersensing		
										A0712	Pacing Problem		
										A071202	Inaccurate Synchronization		
						CT2985	Active	E060109	Tachycardia	A070910	Undersensing		
										A0712	Pacing Problem		
										A071202	Inaccurate Synchronization		
								F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A070910	Undersensing		
										A0712	Pacing Problem		
										A071202	Inaccurate Synchronization		
								F15	Recognized Device or Procedural Complication	A070910	Undersensing		
										A0712	Pacing Problem		
										A071202	Inaccurate Synchronization		
						CT3001	Implantable metal	E060109	Tachycardia	A070910	Undersensing		
										A0712	Pacing Problem		
										A071202	Inaccurate Synchronization		
								F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A070910	Undersensing		
										A0712	Pacing Problem		
										A071202	Inaccurate Synchronization		
								F15	Recognized Device or Procedural Complication	A070910	Undersensing		
										A0712	Pacing Problem		
										A071202	Inaccurate Synchronization		
						CT3002	Implantable synthetic polymer	E060109	Tachycardia	A070910	Undersensing		
										A0712	Pacing Problem		
										A071202	Inaccurate Synchronization		
								F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A070910	Undersensing		
										A0712	Pacing Problem		
										A071202	Inaccurate Synchronization		
								F15	Recognized Device or Procedural Complication	A070910	Undersensing		
										A0712	Pacing Problem		
										A071202	Inaccurate Synchronization		
						CT3006	Implantable inorganic material	E060109	Tachycardia	A070910	Undersensing		
										A0712	Pacing Problem		
										A071202	Inaccurate Synchronization		
								F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A070910	Undersensing		
										A0712	Pacing Problem		
										A071202	Inaccurate Synchronization		
								F15	Recognized Device or Procedural Complication	A070910	Undersensing		
										A0712	Pacing Problem		
										A071202	Inaccurate Synchronization		
						CT314	Electrically powered	E060109	Tachycardia	A070910	Undersensing		
										A0712	Pacing Problem		
										A071202	Inaccurate Synchronization		
								F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A070910	Undersensing		

									A0712	Pacing Problem		
									A071202	Inaccurate Synchronization		
							F15	Recognized Device or Procedural Complication	A070910	Undersensing		
									A0712	Pacing Problem		
									A071202	Inaccurate Synchronization		
					CT321	Long-term surgical invasive	E060109	Tachycardia	A070910	Undersensing		
									A0712	Pacing Problem		
									A071202	Inaccurate Synchronization		
							F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A070910	Undersensing		
									A0712	Pacing Problem		
									A071202	Inaccurate Synchronization		
							F15	Recognized Device or Procedural Complication	A070910	Undersensing		
									A0712	Pacing Problem		
									A071202	Inaccurate Synchronization		
					CT449	Generators and associated devices	E060109	Tachycardia	A070910	Undersensing		
									A0712	Pacing Problem		
									A071202	Inaccurate Synchronization		
							F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A070910	Undersensing		
									A0712	Pacing Problem		
									A071202	Inaccurate Synchronization		
							F15	Recognized Device or Procedural Complication	A070910	Undersensing		
									A0712	Pacing Problem		
									A071202	Inaccurate Synchronization		
					CT981	Single use	E060109	Tachycardia	A070910	Undersensing		
									A0712	Pacing Problem		
									A071202	Inaccurate Synchronization		
							F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A070910	Undersensing		
									A0712	Pacing Problem		
									A071202	Inaccurate Synchronization		
							F15	Recognized Device or Procedural Complication	A070910	Undersensing		
									A0712	Pacing Problem		
									A071202	Inaccurate Synchronization		
					CT983	Surgical invasive	E060109	Tachycardia	A070910	Undersensing		
									A0712	Pacing Problem		
									A071202	Inaccurate Synchronization		
							F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A070910	Undersensing		
									A0712	Pacing Problem		
									A071202	Inaccurate Synchronization		
							F15	Recognized Device or Procedural Complication	A070910	Undersensing		
									A0712	Pacing Problem		
									A071202	Inaccurate Synchronization		
13953	2023	All other reportable incidents	47263	Cardiac resynchronization therapy implantable pacemaker	CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	23/07/2024	23/07/2024
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	23/07/2024	23/07/2024
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	23/07/2024	23/07/2024
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	23/07/2024	23/07/2024
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	23/07/2024	23/07/2024
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	23/07/2024	23/07/2024
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	23/07/2024	23/07/2024
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	23/07/2024	23/07/2024
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	23/07/2024	23/07/2024
13954	2023		47270	Cardiac resynchronization therapy implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	07/09/2023	07/09/2023
							F1901	Additional Surgery	A070504	Premature Discharge of Battery	07/09/2023	07/09/2023
13955	2023	All other reportable incidents	47267	Single-chamber implantable pacemaker, rate-responsive	CT141	Electrodes/Leads and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	07/06/2023	07/06/2023
							F1901	Additional Surgery	A070504	Premature Discharge of Battery	07/06/2023	07/06/2023
					CT145	Cardiovascular devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	07/06/2023	07/06/2023
							F1901	Additional Surgery	A070504	Premature Discharge of Battery	07/06/2023	07/06/2023
					CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	07/06/2023	07/06/2023
							F1901	Additional Surgery	A070504	Premature Discharge of Battery	07/06/2023	07/06/2023
					CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	07/06/2023	07/06/2023
							F1901	Additional Surgery	A070504	Premature Discharge of Battery	07/06/2023	07/06/2023
					CT233	Surgical	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	07/06/2023	07/06/2023
							F1901	Additional Surgery	A070504	Premature Discharge of Battery	07/06/2023	07/06/2023
					CT2406	Implantable	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	07/06/2023	07/06/2023
							F1901	Additional Surgery	A070504	Premature Discharge of Battery	07/06/2023	07/06/2023
					CT2407	Active implantable	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	07/06/2023	07/06/2023
							F1901	Additional Surgery	A070504	Premature Discharge of Battery	07/06/2023	07/06/2023
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	07/06/2023	07/06/2023
							F1901	Additional Surgery	A070504	Premature Discharge of Battery	07/06/2023	07/06/2023
					CT2727	Active implant devices/ancillaries	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	07/06/2023	07/06/2023
							F1901	Additional Surgery	A070504	Premature Discharge of Battery	07/06/2023	07/06/2023
					CT2750	Implantable cardiac electrical stimulators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	07/06/2023	07/06/2023
							F1901	Additional Surgery	A070504	Premature Discharge of Battery	07/06/2023	07/06/2023
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	07/06/2023	07/06/2023
							F1901	Additional Surgery	A070504	Premature Discharge of Battery	07/06/2023	07/06/2023
					CT2985	Active	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	07/06/2023	07/06/2023
							F1901	Additional Surgery	A070504	Premature Discharge of Battery	07/06/2023	07/06/2023
					CT3001	Implantable metal	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	07/06/2023	07/06/2023
							F1901	Additional Surgery	A070504	Premature Discharge of Battery	07/06/2023	07/06/2023
					CT3002	Implantable synthetic polymer	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	07/06/2023	07/06/2023
							F1901	Additional Surgery	A070504	Premature Discharge of Battery	07/06/2023	07/06/2023
					CT3006	Implantable inorganic material	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	07/06/2023	07/06/2023
							F1901	Additional Surgery	A070504	Premature Discharge of Battery	07/06/2023	07/06/2023
					CT314	Electrically powered	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	07/06/2023	07/06/2023
							F1901	Additional Surgery	A070504	Premature Discharge of Battery	07/06/2023	07/06/2023
					CT321	Long-term surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	07/06/2023	07/06/2023
							F1901	Additional Surgery	A070504	Premature Discharge of Battery	07/06/2023	07/06/2023
					CT449	Generators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	07/06/2023	07/06/2023
							F1901	Additional Surgery	A070504	Premature Discharge of Battery	07/06/2023	07/06/2023

					CT981	Single use	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	07/06/2023	07/06/2023
							F1901	Additional Surgery	A070504	Premature Discharge of Battery	07/06/2023	07/06/2023
					CT983	Surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	07/06/2023	07/06/2023
							F1901	Additional Surgery	A070504	Premature Discharge of Battery	07/06/2023	07/06/2023
13956	2023	All other reportable incidents	17882	Physiologic-monitoring defibrillation system	CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A0719	Unexpected Shutdown		
							F27	Problem Identified During Non-Clinical Procedure	A0719	Unexpected Shutdown		
13957	2023	All other reportable incidents	35852	Single-chamber implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	26/09/2023	26/09/2023
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	26/09/2023	26/09/2023
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	26/09/2023	26/09/2023
13958	2023	All other reportable incidents	47265	Dual-chamber implantable pacemaker, rate-responsive	CT141	Electrodes/Leads and associated devices	E0601	Arrhythmia	A0711	Interrogation Problem	24/08/2023	24/08/2023
									A0712	Pacing Problem	24/08/2023	24/08/2023
									A13	Communication or Transmission Problem	24/08/2023	24/08/2023
									A2302	Use of Incorrect Control/Treatment Settings	24/08/2023	24/08/2023
							F1901	Additional Surgery	A0711	Interrogation Problem	24/08/2023	24/08/2023
									A0712	Pacing Problem	24/08/2023	24/08/2023
									A13	Communication or Transmission Problem	24/08/2023	24/08/2023
									A2302	Use of Incorrect Control/Treatment Settings	24/08/2023	24/08/2023
					CT145	Cardiovascular devices	E0601	Arrhythmia	A0711	Interrogation Problem	24/08/2023	24/08/2023
									A0712	Pacing Problem	24/08/2023	24/08/2023
									A13	Communication or Transmission Problem	24/08/2023	24/08/2023
									A2302	Use of Incorrect Control/Treatment Settings	24/08/2023	24/08/2023
							F1901	Additional Surgery	A0711	Interrogation Problem	24/08/2023	24/08/2023
									A0712	Pacing Problem	24/08/2023	24/08/2023
									A13	Communication or Transmission Problem	24/08/2023	24/08/2023
									A2302	Use of Incorrect Control/Treatment Settings	24/08/2023	24/08/2023
					CT1684	Cardiac pacemakers and associated devices	E0601	Arrhythmia	A0711	Interrogation Problem	24/08/2023	24/08/2023
									A0712	Pacing Problem	24/08/2023	24/08/2023
									A13	Communication or Transmission Problem	24/08/2023	24/08/2023
									A2302	Use of Incorrect Control/Treatment Settings	24/08/2023	24/08/2023
							F1901	Additional Surgery	A0711	Interrogation Problem	24/08/2023	24/08/2023
									A0712	Pacing Problem	24/08/2023	24/08/2023
									A13	Communication or Transmission Problem	24/08/2023	24/08/2023
									A2302	Use of Incorrect Control/Treatment Settings	24/08/2023	24/08/2023
					CT1685	Cardiac pacemakers	E0601	Arrhythmia	A0711	Interrogation Problem	24/08/2023	24/08/2023
									A0712	Pacing Problem	24/08/2023	24/08/2023
									A13	Communication or Transmission Problem	24/08/2023	24/08/2023
									A2302	Use of Incorrect Control/Treatment Settings	24/08/2023	24/08/2023
							F1901	Additional Surgery	A0711	Interrogation Problem	24/08/2023	24/08/2023
									A0712	Pacing Problem	24/08/2023	24/08/2023
									A13	Communication or Transmission Problem	24/08/2023	24/08/2023
									A2302	Use of Incorrect Control/Treatment Settings	24/08/2023	24/08/2023
					CT233	Surgical	E0601	Arrhythmia	A0711	Interrogation Problem	24/08/2023	24/08/2023
									A0712	Pacing Problem	24/08/2023	24/08/2023
									A13	Communication or Transmission Problem	24/08/2023	24/08/2023
									A2302	Use of Incorrect Control/Treatment Settings	24/08/2023	24/08/2023
							F1901	Additional Surgery	A0711	Interrogation Problem	24/08/2023	24/08/2023
									A0712	Pacing Problem	24/08/2023	24/08/2023
									A13	Communication or Transmission Problem	24/08/2023	24/08/2023
									A2302	Use of Incorrect Control/Treatment Settings	24/08/2023	24/08/2023
					CT2406	Implantable	E0601	Arrhythmia	A0711	Interrogation Problem	24/08/2023	24/08/2023
									A0712	Pacing Problem	24/08/2023	24/08/2023
									A13	Communication or Transmission Problem	24/08/2023	24/08/2023
									A2302	Use of Incorrect Control/Treatment Settings	24/08/2023	24/08/2023
							F1901	Additional Surgery	A0711	Interrogation Problem	24/08/2023	24/08/2023
									A0712	Pacing Problem	24/08/2023	24/08/2023
									A13	Communication or Transmission Problem	24/08/2023	24/08/2023
									A2302	Use of Incorrect Control/Treatment Settings	24/08/2023	24/08/2023
					CT2407	Active implantable	E0601	Arrhythmia	A0711	Interrogation Problem	24/08/2023	24/08/2023
									A0712	Pacing Problem	24/08/2023	24/08/2023
									A13	Communication or Transmission Problem	24/08/2023	24/08/2023
									A2302	Use of Incorrect Control/Treatment Settings	24/08/2023	24/08/2023
							F1901	Additional Surgery	A0711	Interrogation Problem	24/08/2023	24/08/2023
									A0712	Pacing Problem	24/08/2023	24/08/2023
									A13	Communication or Transmission Problem	24/08/2023	24/08/2023
									A2302	Use of Incorrect Control/Treatment Settings	24/08/2023	24/08/2023
					CT257	Implantable cardiac pacemakers	E0601	Arrhythmia	A0711	Interrogation Problem	24/08/2023	24/08/2023
									A0712	Pacing Problem	24/08/2023	24/08/2023
									A13	Communication or Transmission Problem	24/08/2023	24/08/2023
									A2302	Use of Incorrect Control/Treatment Settings	24/08/2023	24/08/2023
							F1901	Additional Surgery	A0711	Interrogation Problem	24/08/2023	24/08/2023
									A0712	Pacing Problem	24/08/2023	24/08/2023
									A13	Communication or Transmission Problem	24/08/2023	24/08/2023
									A2302	Use of Incorrect Control/Treatment Settings	24/08/2023	24/08/2023
					CT2727	Active implant devices/ancillaries	E0601	Arrhythmia	A0711	Interrogation Problem	24/08/2023	24/08/2023
									A0712	Pacing Problem	24/08/2023	24/08/2023
									A13	Communication or Transmission Problem	24/08/2023	24/08/2023
									A2302	Use of Incorrect Control/Treatment Settings	24/08/2023	24/08/2023
							F1901	Additional Surgery	A0711	Interrogation Problem	24/08/2023	24/08/2023
									A0712	Pacing Problem	24/08/2023	24/08/2023
									A13	Communication or Transmission Problem	24/08/2023	24/08/2023
									A2302	Use of Incorrect Control/Treatment Settings	24/08/2023	24/08/2023
					CT2750	Implantable cardiac electrical stimulators and associated devices	E0601	Arrhythmia	A0711	Interrogation Problem	24/08/2023	24/08/2023
									A0712	Pacing Problem	24/08/2023	24/08/2023
									A13	Communication or Transmission Problem	24/08/2023	24/08/2023
									A2302	Use of Incorrect Control/Treatment Settings	24/08/2023	24/08/2023
							F1901	Additional Surgery	A0711	Interrogation Problem	24/08/2023	24/08/2023
									A0712	Pacing Problem	24/08/2023	24/08/2023
									A13	Communication or Transmission Problem	24/08/2023	24/08/2023

									A2302	Use of Incorrect Control/Treatment Settings	24/08/2023	24/08/2023
					CT2866	Implantable cardiac electrical stimulators	E0601	Arrhythmia	A0711	Interrogation Problem	24/08/2023	24/08/2023
									A0712	Pacing Problem	24/08/2023	24/08/2023
									A13	Communication or Transmission Problem	24/08/2023	24/08/2023
									A2302	Use of Incorrect Control/Treatment Settings	24/08/2023	24/08/2023
							F1901	Additional Surgery	A0711	Interrogation Problem	24/08/2023	24/08/2023
									A0712	Pacing Problem	24/08/2023	24/08/2023
									A13	Communication or Transmission Problem	24/08/2023	24/08/2023
									A2302	Use of Incorrect Control/Treatment Settings	24/08/2023	24/08/2023
					CT2985	Active	E0601	Arrhythmia	A0711	Interrogation Problem	24/08/2023	24/08/2023
									A0712	Pacing Problem	24/08/2023	24/08/2023
									A13	Communication or Transmission Problem	24/08/2023	24/08/2023
									A2302	Use of Incorrect Control/Treatment Settings	24/08/2023	24/08/2023
							F1901	Additional Surgery	A0711	Interrogation Problem	24/08/2023	24/08/2023
									A0712	Pacing Problem	24/08/2023	24/08/2023
									A13	Communication or Transmission Problem	24/08/2023	24/08/2023
									A2302	Use of Incorrect Control/Treatment Settings	24/08/2023	24/08/2023
					CT3001	Implantable metal	E0601	Arrhythmia	A0711	Interrogation Problem	24/08/2023	24/08/2023
									A0712	Pacing Problem	24/08/2023	24/08/2023
									A13	Communication or Transmission Problem	24/08/2023	24/08/2023
									A2302	Use of Incorrect Control/Treatment Settings	24/08/2023	24/08/2023
							F1901	Additional Surgery	A0711	Interrogation Problem	24/08/2023	24/08/2023
									A0712	Pacing Problem	24/08/2023	24/08/2023
									A13	Communication or Transmission Problem	24/08/2023	24/08/2023
									A2302	Use of Incorrect Control/Treatment Settings	24/08/2023	24/08/2023
					CT3002	Implantable synthetic polymer	E0601	Arrhythmia	A0711	Interrogation Problem	24/08/2023	24/08/2023
									A0712	Pacing Problem	24/08/2023	24/08/2023
									A13	Communication or Transmission Problem	24/08/2023	24/08/2023
									A2302	Use of Incorrect Control/Treatment Settings	24/08/2023	24/08/2023
							F1901	Additional Surgery	A0711	Interrogation Problem	24/08/2023	24/08/2023
									A0712	Pacing Problem	24/08/2023	24/08/2023
									A13	Communication or Transmission Problem	24/08/2023	24/08/2023
									A2302	Use of Incorrect Control/Treatment Settings	24/08/2023	24/08/2023
					CT3006	Implantable inorganic material	E0601	Arrhythmia	A0711	Interrogation Problem	24/08/2023	24/08/2023
									A0712	Pacing Problem	24/08/2023	24/08/2023
									A13	Communication or Transmission Problem	24/08/2023	24/08/2023
									A2302	Use of Incorrect Control/Treatment Settings	24/08/2023	24/08/2023
							F1901	Additional Surgery	A0711	Interrogation Problem	24/08/2023	24/08/2023
									A0712	Pacing Problem	24/08/2023	24/08/2023
									A13	Communication or Transmission Problem	24/08/2023	24/08/2023
									A2302	Use of Incorrect Control/Treatment Settings	24/08/2023	24/08/2023
					CT314	Electrically powered	E0601	Arrhythmia	A0711	Interrogation Problem	24/08/2023	24/08/2023
									A0712	Pacing Problem	24/08/2023	24/08/2023
									A13	Communication or Transmission Problem	24/08/2023	24/08/2023
									A2302	Use of Incorrect Control/Treatment Settings	24/08/2023	24/08/2023
							F1901	Additional Surgery	A0711	Interrogation Problem	24/08/2023	24/08/2023
									A0712	Pacing Problem	24/08/2023	24/08/2023
									A13	Communication or Transmission Problem	24/08/2023	24/08/2023
									A2302	Use of Incorrect Control/Treatment Settings	24/08/2023	24/08/2023
					CT321	Long-term surgical invasive	E0601	Arrhythmia	A0711	Interrogation Problem	24/08/2023	24/08/2023
									A0712	Pacing Problem	24/08/2023	24/08/2023
									A13	Communication or Transmission Problem	24/08/2023	24/08/2023
									A2302	Use of Incorrect Control/Treatment Settings	24/08/2023	24/08/2023
							F1901	Additional Surgery	A0711	Interrogation Problem	24/08/2023	24/08/2023
									A0712	Pacing Problem	24/08/2023	24/08/2023
									A13	Communication or Transmission Problem	24/08/2023	24/08/2023
									A2302	Use of Incorrect Control/Treatment Settings	24/08/2023	24/08/2023
					CT449	Generators and associated devices	E0601	Arrhythmia	A0711	Interrogation Problem	24/08/2023	24/08/2023
									A0712	Pacing Problem	24/08/2023	24/08/2023
									A13	Communication or Transmission Problem	24/08/2023	24/08/2023
									A2302	Use of Incorrect Control/Treatment Settings	24/08/2023	24/08/2023
							F1901	Additional Surgery	A0711	Interrogation Problem	24/08/2023	24/08/2023
									A0712	Pacing Problem	24/08/2023	24/08/2023
									A13	Communication or Transmission Problem	24/08/2023	24/08/2023
									A2302	Use of Incorrect Control/Treatment Settings	24/08/2023	24/08/2023
					CT981	Single use	E0601	Arrhythmia	A0711	Interrogation Problem	24/08/2023	24/08/2023
									A0712	Pacing Problem	24/08/2023	24/08/2023
									A13	Communication or Transmission Problem	24/08/2023	24/08/2023
									A2302	Use of Incorrect Control/Treatment Settings	24/08/2023	24/08/2023
							F1901	Additional Surgery	A0711	Interrogation Problem	24/08/2023	24/08/2023
									A0712	Pacing Problem	24/08/2023	24/08/2023
									A13	Communication or Transmission Problem	24/08/2023	24/08/2023
									A2302	Use of Incorrect Control/Treatment Settings	24/08/2023	24/08/2023
					CT983	Surgical invasive	E0601	Arrhythmia	A0711	Interrogation Problem	24/08/2023	24/08/2023
									A0712	Pacing Problem	24/08/2023	24/08/2023
									A13	Communication or Transmission Problem	24/08/2023	24/08/2023
									A2302	Use of Incorrect Control/Treatment Settings	24/08/2023	24/08/2023
							F1901	Additional Surgery	A0711	Interrogation Problem	24/08/2023	24/08/2023
									A0712	Pacing Problem	24/08/2023	24/08/2023
									A13	Communication or Transmission Problem	24/08/2023	24/08/2023
									A2302	Use of Incorrect Control/Treatment Settings	24/08/2023	24/08/2023
13959	2023		47265	Dual-chamber implantable pacemaker, rate-responsive	CT141	Electrodes/Leads and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A090801	Signal Artifact/Noise	25/08/2023	25/08/2023
							F1901	Additional Surgery	A090801	Signal Artifact/Noise	25/08/2023	25/08/2023
					CT145	Cardiovascular devices	E2403	No Clinical Signs, Symptoms or Conditions	A090801	Signal Artifact/Noise	25/08/2023	25/08/2023
							F1901	Additional Surgery	A090801	Signal Artifact/Noise	25/08/2023	25/08/2023
					CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A090801	Signal Artifact/Noise	25/08/2023	25/08/2023
							F1901	Additional Surgery	A090801	Signal Artifact/Noise	25/08/2023	25/08/2023
					CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A090801	Signal Artifact/Noise	25/08/2023	25/08/2023

							F1901	Additional Surgery	A090801	Signal Artifact/Noise	25/08/2023	25/08/2023
					CT233	Surgical	E2403	No Clinical Signs, Symptoms or Conditions	A090801	Signal Artifact/Noise	25/08/2023	25/08/2023
							F1901	Additional Surgery	A090801	Signal Artifact/Noise	25/08/2023	25/08/2023
					CT2406	Implantable	E2403	No Clinical Signs, Symptoms or Conditions	A090801	Signal Artifact/Noise	25/08/2023	25/08/2023
							F1901	Additional Surgery	A090801	Signal Artifact/Noise	25/08/2023	25/08/2023
					CT2407	Active implantable	E2403	No Clinical Signs, Symptoms or Conditions	A090801	Signal Artifact/Noise	25/08/2023	25/08/2023
							F1901	Additional Surgery	A090801	Signal Artifact/Noise	25/08/2023	25/08/2023
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A090801	Signal Artifact/Noise	25/08/2023	25/08/2023
							F1901	Additional Surgery	A090801	Signal Artifact/Noise	25/08/2023	25/08/2023
					CT2727	Active implant devices/ancillaries	E2403	No Clinical Signs, Symptoms or Conditions	A090801	Signal Artifact/Noise	25/08/2023	25/08/2023
							F1901	Additional Surgery	A090801	Signal Artifact/Noise	25/08/2023	25/08/2023
					CT2750	Implantable cardiac electrical stimulators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A090801	Signal Artifact/Noise	25/08/2023	25/08/2023
							F1901	Additional Surgery	A090801	Signal Artifact/Noise	25/08/2023	25/08/2023
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A090801	Signal Artifact/Noise	25/08/2023	25/08/2023
							F1901	Additional Surgery	A090801	Signal Artifact/Noise	25/08/2023	25/08/2023
					CT2985	Active	E2403	No Clinical Signs, Symptoms or Conditions	A090801	Signal Artifact/Noise	25/08/2023	25/08/2023
							F1901	Additional Surgery	A090801	Signal Artifact/Noise	25/08/2023	25/08/2023
					CT3001	Implantable metal	E2403	No Clinical Signs, Symptoms or Conditions	A090801	Signal Artifact/Noise	25/08/2023	25/08/2023
							F1901	Additional Surgery	A090801	Signal Artifact/Noise	25/08/2023	25/08/2023
					CT3002	Implantable synthetic polymer	E2403	No Clinical Signs, Symptoms or Conditions	A090801	Signal Artifact/Noise	25/08/2023	25/08/2023
							F1901	Additional Surgery	A090801	Signal Artifact/Noise	25/08/2023	25/08/2023
					CT3006	Implantable inorganic material	E2403	No Clinical Signs, Symptoms or Conditions	A090801	Signal Artifact/Noise	25/08/2023	25/08/2023
							F1901	Additional Surgery	A090801	Signal Artifact/Noise	25/08/2023	25/08/2023
					CT314	Electrically powered	E2403	No Clinical Signs, Symptoms or Conditions	A090801	Signal Artifact/Noise	25/08/2023	25/08/2023
							F1901	Additional Surgery	A090801	Signal Artifact/Noise	25/08/2023	25/08/2023
					CT321	Long-term surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A090801	Signal Artifact/Noise	25/08/2023	25/08/2023
							F1901	Additional Surgery	A090801	Signal Artifact/Noise	25/08/2023	25/08/2023
					CT449	Generators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A090801	Signal Artifact/Noise	25/08/2023	25/08/2023
							F1901	Additional Surgery	A090801	Signal Artifact/Noise	25/08/2023	25/08/2023
					CT981	Single use	E2403	No Clinical Signs, Symptoms or Conditions	A090801	Signal Artifact/Noise	25/08/2023	25/08/2023
							F1901	Additional Surgery	A090801	Signal Artifact/Noise	25/08/2023	25/08/2023
					CT983	Surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A090801	Signal Artifact/Noise	25/08/2023	25/08/2023
							F1901	Additional Surgery	A090801	Signal Artifact/Noise	25/08/2023	25/08/2023
13960	2023	All other reportable incidents	17882	Physiologic-monitoring defibrillation system	CT1684	Cardiac pacemakers and associated devices	E0601	Arrhythmia	A07	Electrical /Electronic Property Problem		
							F05	Delay to Treatment/ Therapy	A07	Electrical /Electronic Property Problem		
13961	2023	All other reportable incidents	47265	Dual-chamber implantable pacemaker, rate-responsive	CT141	Electrodes/Leads and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A09	Output Problem	29/08/2023	29/08/2023
							F08	Hospitalization or Prolonged Hospitalization	A09	Output Problem	29/08/2023	29/08/2023
					CT145	Cardiovascular devices	E2403	No Clinical Signs, Symptoms or Conditions	A09	Output Problem	29/08/2023	29/08/2023
							F08	Hospitalization or Prolonged Hospitalization	A09	Output Problem	29/08/2023	29/08/2023
					CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A09	Output Problem	29/08/2023	29/08/2023
							F08	Hospitalization or Prolonged Hospitalization	A09	Output Problem	29/08/2023	29/08/2023
					CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A09	Output Problem	29/08/2023	29/08/2023
							F08	Hospitalization or Prolonged Hospitalization	A09	Output Problem	29/08/2023	29/08/2023
					CT233	Surgical	E2403	No Clinical Signs, Symptoms or Conditions	A09	Output Problem	29/08/2023	29/08/2023
							F08	Hospitalization or Prolonged Hospitalization	A09	Output Problem	29/08/2023	29/08/2023
					CT2406	Implantable	E2403	No Clinical Signs, Symptoms or Conditions	A09	Output Problem	29/08/2023	29/08/2023
							F08	Hospitalization or Prolonged Hospitalization	A09	Output Problem	29/08/2023	29/08/2023
					CT2407	Active implantable	E2403	No Clinical Signs, Symptoms or Conditions	A09	Output Problem	29/08/2023	29/08/2023
							F08	Hospitalization or Prolonged Hospitalization	A09	Output Problem	29/08/2023	29/08/2023
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A09	Output Problem	29/08/2023	29/08/2023
							F08	Hospitalization or Prolonged Hospitalization	A09	Output Problem	29/08/2023	29/08/2023
					CT2727	Active implant devices/ancillaries	E2403	No Clinical Signs, Symptoms or Conditions	A09	Output Problem	29/08/2023	29/08/2023
							F08	Hospitalization or Prolonged Hospitalization	A09	Output Problem	29/08/2023	29/08/2023
					CT2750	Implantable cardiac electrical stimulators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A09	Output Problem	29/08/2023	29/08/2023
							F08	Hospitalization or Prolonged Hospitalization	A09	Output Problem	29/08/2023	29/08/2023
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A09	Output Problem	29/08/2023	29/08/2023
							F08	Hospitalization or Prolonged Hospitalization	A09	Output Problem	29/08/2023	29/08/2023
					CT2985	Active	E2403	No Clinical Signs, Symptoms or Conditions	A09	Output Problem	29/08/2023	29/08/2023
							F08	Hospitalization or Prolonged Hospitalization	A09	Output Problem	29/08/2023	29/08/2023
					CT3001	Implantable metal	E2403	No Clinical Signs, Symptoms or Conditions	A09	Output Problem	29/08/2023	29/08/2023
							F08	Hospitalization or Prolonged Hospitalization	A09	Output Problem	29/08/2023	29/08/2023
					CT3002	Implantable synthetic polymer	E2403	No Clinical Signs, Symptoms or Conditions	A09	Output Problem	29/08/2023	29/08/2023
							F08	Hospitalization or Prolonged Hospitalization	A09	Output Problem	29/08/2023	29/08/2023
					CT3006	Implantable inorganic material	E2403	No Clinical Signs, Symptoms or Conditions	A09	Output Problem	29/08/2023	29/08/2023
							F08	Hospitalization or Prolonged Hospitalization	A09	Output Problem	29/08/2023	29/08/2023
					CT314	Electrically powered	E2403	No Clinical Signs, Symptoms or Conditions	A09	Output Problem	29/08/2023	29/08/2023
							F08	Hospitalization or Prolonged Hospitalization	A09	Output Problem	29/08/2023	29/08/2023
					CT321	Long-term surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A09	Output Problem	29/08/2023	29/08/2023
							F08	Hospitalization or Prolonged Hospitalization	A09	Output Problem	29/08/2023	29/08/2023
					CT449	Generators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A09	Output Problem	29/08/2023	29/08/2023
							F08	Hospitalization or Prolonged Hospitalization	A09	Output Problem	29/08/2023	29/08/2023
					CT981	Single use	E2403	No Clinical Signs, Symptoms or Conditions	A09	Output Problem	29/08/2023	29/08/2023
							F08	Hospitalization or Prolonged Hospitalization	A09	Output Problem	29/08/2023	29/08/2023
					CT983	Surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A09	Output Problem	29/08/2023	29/08/2023
							F08	Hospitalization or Prolonged Hospitalization	A09	Output Problem	29/08/2023	29/08/2023
13962	2023	All other reportable incidents	47265	Dual-chamber implantable pacemaker, rate-responsive	CT141	Electrodes/Leads and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A0908	Incorrect, Inadequate or Imprecise Result or Readings		
							F26	No Health Consequences or Impact	A0908	Incorrect, Inadequate or Imprecise Result or Readings		
					CT145	Cardiovascular devices	E2403	No Clinical Signs, Symptoms or Conditions	A0908	Incorrect, Inadequate or Imprecise Result or Readings		
							F26	No Health Consequences or Impact	A0908	Incorrect, Inadequate or Imprecise Result or Readings		
					CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A0908	Incorrect, Inadequate or Imprecise Result or Readings		
							F26	No Health Consequences or Impact	A0908	Incorrect, Inadequate or Imprecise Result or Readings		
					CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A0908	Incorrect, Inadequate or Imprecise Result or Readings		

							F26	No Health Consequences or Impact	A0908	Incorrect, Inadequate or Imprecise Result or Readings		
					CT233	Surgical	E2403	No Clinical Signs, Symptoms or Conditions	A0908	Incorrect, Inadequate or Imprecise Result or Readings		
							F26	No Health Consequences or Impact	A0908	Incorrect, Inadequate or Imprecise Result or Readings		
					CT2406	Implantable	E2403	No Clinical Signs, Symptoms or Conditions	A0908	Incorrect, Inadequate or Imprecise Result or Readings		
							F26	No Health Consequences or Impact	A0908	Incorrect, Inadequate or Imprecise Result or Readings		
					CT2407	Active implantable	E2403	No Clinical Signs, Symptoms or Conditions	A0908	Incorrect, Inadequate or Imprecise Result or Readings		
							F26	No Health Consequences or Impact	A0908	Incorrect, Inadequate or Imprecise Result or Readings		
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A0908	Incorrect, Inadequate or Imprecise Result or Readings		
							F26	No Health Consequences or Impact	A0908	Incorrect, Inadequate or Imprecise Result or Readings		
					CT2727	Active implant devices/ancillaries	E2403	No Clinical Signs, Symptoms or Conditions	A0908	Incorrect, Inadequate or Imprecise Result or Readings		
							F26	No Health Consequences or Impact	A0908	Incorrect, Inadequate or Imprecise Result or Readings		
					CT2750	Implantable cardiac electrical stimulators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A0908	Incorrect, Inadequate or Imprecise Result or Readings		
							F26	No Health Consequences or Impact	A0908	Incorrect, Inadequate or Imprecise Result or Readings		
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A0908	Incorrect, Inadequate or Imprecise Result or Readings		
							F26	No Health Consequences or Impact	A0908	Incorrect, Inadequate or Imprecise Result or Readings		
					CT2985	Active	E2403	No Clinical Signs, Symptoms or Conditions	A0908	Incorrect, Inadequate or Imprecise Result or Readings		
							F26	No Health Consequences or Impact	A0908	Incorrect, Inadequate or Imprecise Result or Readings		
					CT3001	Implantable metal	E2403	No Clinical Signs, Symptoms or Conditions	A0908	Incorrect, Inadequate or Imprecise Result or Readings		
							F26	No Health Consequences or Impact	A0908	Incorrect, Inadequate or Imprecise Result or Readings		
					CT3002	Implantable synthetic polymer	E2403	No Clinical Signs, Symptoms or Conditions	A0908	Incorrect, Inadequate or Imprecise Result or Readings		
							F26	No Health Consequences or Impact	A0908	Incorrect, Inadequate or Imprecise Result or Readings		
					CT3006	Implantable inorganic material	E2403	No Clinical Signs, Symptoms or Conditions	A0908	Incorrect, Inadequate or Imprecise Result or Readings		
							F26	No Health Consequences or Impact	A0908	Incorrect, Inadequate or Imprecise Result or Readings		
					CT314	Electrically powered	E2403	No Clinical Signs, Symptoms or Conditions	A0908	Incorrect, Inadequate or Imprecise Result or Readings		
							F26	No Health Consequences or Impact	A0908	Incorrect, Inadequate or Imprecise Result or Readings		
					CT321	Long-term surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A0908	Incorrect, Inadequate or Imprecise Result or Readings		
							F26	No Health Consequences or Impact	A0908	Incorrect, Inadequate or Imprecise Result or Readings		
					CT449	Generators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A0908	Incorrect, Inadequate or Imprecise Result or Readings		
							F26	No Health Consequences or Impact	A0908	Incorrect, Inadequate or Imprecise Result or Readings		
					CT981	Single use	E2403	No Clinical Signs, Symptoms or Conditions	A0908	Incorrect, Inadequate or Imprecise Result or Readings		
							F26	No Health Consequences or Impact	A0908	Incorrect, Inadequate or Imprecise Result or Readings		
					CT983	Surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A0908	Incorrect, Inadequate or Imprecise Result or Readings		
							F26	No Health Consequences or Impact	A0908	Incorrect, Inadequate or Imprecise Result or Readings		
13963	2023	All other reportable incidents	36241	Pacemaker/defibrillator lead	CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070102	High Capture Threshold		
									A070909	Oversensing		
									A090801	Signal Artifact/Noise		
							F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A070102	High Capture Threshold		
									A070909	Oversensing		
									A090801	Signal Artifact/Noise		
13964	2023	All other reportable incidents	47270	Cardiac resynchronization therapy implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
									A0908	Incorrect, Inadequate or Imprecise Result or Readings		
							F24	Insufficient Health Impact Information	A070504	Premature Discharge of Battery		
									A0908	Incorrect, Inadequate or Imprecise Result or Readings		
13965	2023	All other reportable incidents	47267	Single-chamber implantable pacemaker, rate-responsive	CT141	Electrodes/Leads and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT145	Cardiovascular devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT233	Surgical	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT2406	Implantable	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT2407	Active implantable	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT2727	Active implant devices/ancillaries	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		

					CT2750	Implantable cardiac electrical stimulators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT2985	Active	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT3001	Implantable metal	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT3002	Implantable synthetic polymer	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT3006	Implantable inorganic material	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT314	Electrically powered	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT321	Long-term surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT449	Generators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT981	Single use	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT983	Surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
13966	2023	All other reportable incidents	47270	Cardiac resynchronization therapy implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E0601	Arrhythmia	A071301	Failure to Deliver Shock/Stimulation		
							F12	Serious Injury/ Illness/ Impairment	A071301	Failure to Deliver Shock/Stimulation		
13967	2023	All other reportable incidents	47265	Dual-chamber implantable pacemaker, rate-responsive	CT141	Electrodes/Leads and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070909	Oversensing		
									A072201	High Impedance		
									A090801	Signal Artifact/Noise		
							F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A070909	Oversensing		
									A072201	High Impedance		
									A090801	Signal Artifact/Noise		
					CT145	Cardiovascular devices	E2403	No Clinical Signs, Symptoms or Conditions	A070909	Oversensing		
									A072201	High Impedance		
									A090801	Signal Artifact/Noise		
							F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A070909	Oversensing		
									A072201	High Impedance		
									A090801	Signal Artifact/Noise		
					CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070909	Oversensing		
									A072201	High Impedance		
									A090801	Signal Artifact/Noise		
							F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A070909	Oversensing		
									A072201	High Impedance		
									A090801	Signal Artifact/Noise		
					CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A070909	Oversensing		
									A072201	High Impedance		
									A090801	Signal Artifact/Noise		
							F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A070909	Oversensing		
									A072201	High Impedance		
									A090801	Signal Artifact/Noise		
					CT233	Surgical	E2403	No Clinical Signs, Symptoms or Conditions	A070909	Oversensing		
									A072201	High Impedance		
									A090801	Signal Artifact/Noise		
							F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A070909	Oversensing		
									A072201	High Impedance		
									A090801	Signal Artifact/Noise		
					CT2406	Implantable	E2403	No Clinical Signs, Symptoms or Conditions	A070909	Oversensing		
									A072201	High Impedance		
									A090801	Signal Artifact/Noise		
							F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A070909	Oversensing		
									A072201	High Impedance		
									A090801	Signal Artifact/Noise		
					CT2407	Active implantable	E2403	No Clinical Signs, Symptoms or Conditions	A070909	Oversensing		
									A072201	High Impedance		
									A090801	Signal Artifact/Noise		
							F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A070909	Oversensing		
									A072201	High Impedance		
									A090801	Signal Artifact/Noise		
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A070909	Oversensing		
									A072201	High Impedance		
									A090801	Signal Artifact/Noise		
							F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A070909	Oversensing		
									A072201	High Impedance		
									A090801	Signal Artifact/Noise		
					CT2727	Active implant devices/ancillaries	E2403	No Clinical Signs, Symptoms or Conditions	A070909	Oversensing		
									A072201	High Impedance		
									A090801	Signal Artifact/Noise		
							F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A070909	Oversensing		
									A072201	High Impedance		
									A090801	Signal Artifact/Noise		
					CT2750	Implantable cardiac electrical stimulators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070909	Oversensing		
									A072201	High Impedance		
									A090801	Signal Artifact/Noise		
							F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A070909	Oversensing		
									A072201	High Impedance		
									A090801	Signal Artifact/Noise		
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A070909	Oversensing		
									A072201	High Impedance		
									A090801	Signal Artifact/Noise		
							F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A070909	Oversensing		

									A072201	High Impedance		
									A090801	Signal Artifact/Noise		
					CT2985	Active	E2403	No Clinical Signs, Symptoms or Conditions	A070909	Oversensing		
									A072201	High Impedance		
									A090801	Signal Artifact/Noise		
							F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A070909	Oversensing		
									A072201	High Impedance		
									A090801	Signal Artifact/Noise		
					CT3001	Implantable metal	E2403	No Clinical Signs, Symptoms or Conditions	A070909	Oversensing		
									A072201	High Impedance		
									A090801	Signal Artifact/Noise		
							F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A070909	Oversensing		
									A072201	High Impedance		
									A090801	Signal Artifact/Noise		
					CT3002	Implantable synthetic polymer	E2403	No Clinical Signs, Symptoms or Conditions	A070909	Oversensing		
									A072201	High Impedance		
									A090801	Signal Artifact/Noise		
							F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A070909	Oversensing		
									A072201	High Impedance		
									A090801	Signal Artifact/Noise		
					CT3006	Implantable inorganic material	E2403	No Clinical Signs, Symptoms or Conditions	A070909	Oversensing		
									A072201	High Impedance		
									A090801	Signal Artifact/Noise		
							F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A070909	Oversensing		
									A072201	High Impedance		
									A090801	Signal Artifact/Noise		
					CT314	Electrically powered	E2403	No Clinical Signs, Symptoms or Conditions	A070909	Oversensing		
									A072201	High Impedance		
									A090801	Signal Artifact/Noise		
							F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A070909	Oversensing		
									A072201	High Impedance		
									A090801	Signal Artifact/Noise		
					CT321	Long-term surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A070909	Oversensing		
									A072201	High Impedance		
									A090801	Signal Artifact/Noise		
							F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A070909	Oversensing		
									A072201	High Impedance		
									A090801	Signal Artifact/Noise		
					CT449	Generators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070909	Oversensing		
									A072201	High Impedance		
									A090801	Signal Artifact/Noise		
							F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A070909	Oversensing		
									A072201	High Impedance		
									A090801	Signal Artifact/Noise		
					CT981	Single use	E2403	No Clinical Signs, Symptoms or Conditions	A070909	Oversensing		
									A072201	High Impedance		
									A090801	Signal Artifact/Noise		
							F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A070909	Oversensing		
									A072201	High Impedance		
									A090801	Signal Artifact/Noise		
					CT983	Surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A070909	Oversensing		
									A072201	High Impedance		
									A090801	Signal Artifact/Noise		
							F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A070909	Oversensing		
									A072201	High Impedance		
									A090801	Signal Artifact/Noise		
13968	2023	All other reportable incidents	47263	Cardiac resynchronization therapy implantable pacemaker	CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070102	High Capture Threshold		
									A070909	Oversensing		
									A090801	Signal Artifact/Noise		
							F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A070102	High Capture Threshold		
									A070909	Oversensing		
									A090801	Signal Artifact/Noise		
					CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A070102	High Capture Threshold		
									A070909	Oversensing		
									A090801	Signal Artifact/Noise		
							F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A070102	High Capture Threshold		
									A070909	Oversensing		
									A090801	Signal Artifact/Noise		
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A070102	High Capture Threshold		
									A070909	Oversensing		
									A090801	Signal Artifact/Noise		
							F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A070102	High Capture Threshold		
									A070909	Oversensing		
									A090801	Signal Artifact/Noise		
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A070102	High Capture Threshold		
									A070909	Oversensing		
									A090801	Signal Artifact/Noise		
							F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A070102	High Capture Threshold		
									A070909	Oversensing		
									A090801	Signal Artifact/Noise		
13969	2023	All other reportable incidents	36241	Pacemaker/defibrillator lead	CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070909	Oversensing		
									A072201	High Impedance		
									A090801	Signal Artifact/Noise		
							F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A070909	Oversensing		
									A072201	High Impedance		
									A090801	Signal Artifact/Noise		
13970	2023	All other reportable incidents	35852	Single-chamber implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	11/09/2023	11/09/2023

									A0908	Incorrect, Inadequate or Imprecise Result or Readings	11/09/2023	11/09/2023
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	11/09/2023	11/09/2023
									A0908	Incorrect, Inadequate or Imprecise Result or Readings	11/09/2023	11/09/2023
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	11/09/2023	11/09/2023
									A0908	Incorrect, Inadequate or Imprecise Result or Readings	11/09/2023	11/09/2023
13971	2023	Death	35852	Single-chamber implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E060109	Tachycardia	A070909	Oversensing	22/08/2023	22/08/2023
							E060110	Ventricular Fibrillation	A070909	Oversensing	22/08/2023	22/08/2023
							E210401	Shock from Patient Lead(s)	A070909	Oversensing	22/08/2023	22/08/2023
							F02	Death	A070909	Oversensing	22/08/2023	22/08/2023
13972	2023	All other reportable incidents	63188	Left-ventricular wireless pacing system sensor/transmitter	CT141	Electrodes/Leads and associated devices	F1903	Device Explantation	A0708	Power Problem		
					CT145	Cardiovascular devices	F1903	Device Explantation	A0708	Power Problem		
					CT1577	Wireless monitoring system transmitters	F1903	Device Explantation	A0708	Power Problem		
					CT1684	Cardiac pacemakers and associated devices	F1903	Device Explantation	A0708	Power Problem		
					CT2317	Transmitters/Receivers/Transceivers	F1903	Device Explantation	A0708	Power Problem		
					CT233	Surgical	F1903	Device Explantation	A0708	Power Problem		
					CT2406	Implantable	F1903	Device Explantation	A0708	Power Problem		
					CT2407	Active implantable	F1903	Device Explantation	A0708	Power Problem		
					CT255	Transducers and associated devices	F1903	Device Explantation	A0708	Power Problem		
					CT2726	Physiological monitoring/measuring devices	F1903	Device Explantation	A0708	Power Problem		
					CT2727	Active implant devices/ancillaries	F1903	Device Explantation	A0708	Power Problem		
					CT2750	Implantable cardiac electrical stimulators and associated devices	F1903	Device Explantation	A0708	Power Problem		
					CT2885	Physiological monitoring/measuring system transmitters/receivers/transceivers	F1903	Device Explantation	A0708	Power Problem		
					CT2985	Active	F1903	Device Explantation	A0708	Power Problem		
					CT3001	Implantable metal	F1903	Device Explantation	A0708	Power Problem		
					CT3002	Implantable synthetic polymer	F1903	Device Explantation	A0708	Power Problem		
					CT3006	Implantable inorganic material	F1903	Device Explantation	A0708	Power Problem		
					CT314	Electrically powered	F1903	Device Explantation	A0708	Power Problem		
					CT321	Long-term surgical invasive	F1903	Device Explantation	A0708	Power Problem		
					CT384	Sensors and associated devices	F1903	Device Explantation	A0708	Power Problem		
					CT449	Generators and associated devices	F1903	Device Explantation	A0708	Power Problem		
					CT981	Single use	F1903	Device Explantation	A0708	Power Problem		
					CT983	Surgical invasive	F1903	Device Explantation	A0708	Power Problem		
					CT994	Sensors	F1903	Device Explantation	A0708	Power Problem		
13973	2023	All other reportable incidents			CT1684	Cardiac pacemakers and associated devices			A0708	Power Problem		
13974	2023	All other reportable incidents	47263	Cardiac resynchronization therapy implantable pacemaker	CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A0203	Defective Device	25/05/2023	25/05/2023
							F08	Hospitalization or Prolonged Hospitalization	A0203	Defective Device	25/05/2023	25/05/2023
							F1905	Device Revision or Replacement	A0203	Defective Device	25/05/2023	25/05/2023
					CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A0203	Defective Device	25/05/2023	25/05/2023
							F08	Hospitalization or Prolonged Hospitalization	A0203	Defective Device	25/05/2023	25/05/2023
							F1905	Device Revision or Replacement	A0203	Defective Device	25/05/2023	25/05/2023
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A0203	Defective Device	25/05/2023	25/05/2023
							F08	Hospitalization or Prolonged Hospitalization	A0203	Defective Device	25/05/2023	25/05/2023
							F1905	Device Revision or Replacement	A0203	Defective Device	25/05/2023	25/05/2023
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A0203	Defective Device	25/05/2023	25/05/2023
							F08	Hospitalization or Prolonged Hospitalization	A0203	Defective Device	25/05/2023	25/05/2023
							F1905	Device Revision or Replacement	A0203	Defective Device	25/05/2023	25/05/2023
13975	2023	All other reportable incidents	35223	Endocardial/interventricular septal pacing lead	CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070909	Oversensing		
							F23	Unexpected Medical Intervention	A070909	Oversensing		
13976	2023		47265	Dual-chamber implantable pacemaker, rate-responsive	CT141	Electrodes/Leads and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	18/05/2023	18/05/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	18/05/2023	18/05/2023
					CT145	Cardiovascular devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	18/05/2023	18/05/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	18/05/2023	18/05/2023
					CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	18/05/2023	18/05/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	18/05/2023	18/05/2023
					CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	18/05/2023	18/05/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	18/05/2023	18/05/2023
					CT233	Surgical	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	18/05/2023	18/05/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	18/05/2023	18/05/2023
					CT2406	Implantable	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	18/05/2023	18/05/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	18/05/2023	18/05/2023
					CT2407	Active implantable	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	18/05/2023	18/05/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	18/05/2023	18/05/2023
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	18/05/2023	18/05/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	18/05/2023	18/05/2023
					CT2727	Active implant devices/ancillaries	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	18/05/2023	18/05/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	18/05/2023	18/05/2023
					CT2750	Implantable cardiac electrical stimulators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	18/05/2023	18/05/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	18/05/2023	18/05/2023
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	18/05/2023	18/05/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	18/05/2023	18/05/2023
					CT2985	Active	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	18/05/2023	18/05/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	18/05/2023	18/05/2023
					CT3001	Implantable metal	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	18/05/2023	18/05/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	18/05/2023	18/05/2023
					CT3002	Implantable synthetic polymer	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	18/05/2023	18/05/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	18/05/2023	18/05/2023
					CT3006	Implantable inorganic material	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	18/05/2023	18/05/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	18/05/2023	18/05/2023
					CT314	Electrically powered	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	18/05/2023	18/05/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	18/05/2023	18/05/2023
					CT321	Long-term surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	18/05/2023	18/05/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	18/05/2023	18/05/2023
					CT449	Generators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	18/05/2023	18/05/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	18/05/2023	18/05/2023

					CT981	Single use	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	18/05/2023	18/05/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	18/05/2023	18/05/2023
					CT983	Surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	18/05/2023	18/05/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	18/05/2023	18/05/2023
13977	2023		47265	Dual-chamber implantable pacemaker, rate-responsive	CT141	Electrodes/Leads and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	09/06/2023	09/06/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	09/06/2023	09/06/2023
					CT145	Cardiovascular devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	09/06/2023	09/06/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	09/06/2023	09/06/2023
					CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	09/06/2023	09/06/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	09/06/2023	09/06/2023
					CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	09/06/2023	09/06/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	09/06/2023	09/06/2023
					CT233	Surgical	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	09/06/2023	09/06/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	09/06/2023	09/06/2023
					CT2406	Implantable	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	09/06/2023	09/06/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	09/06/2023	09/06/2023
					CT2407	Active implantable	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	09/06/2023	09/06/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	09/06/2023	09/06/2023
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	09/06/2023	09/06/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	09/06/2023	09/06/2023
					CT2727	Active implant devices/ancillaries	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	09/06/2023	09/06/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	09/06/2023	09/06/2023
					CT2750	Implantable cardiac electrical stimulators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	09/06/2023	09/06/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	09/06/2023	09/06/2023
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	09/06/2023	09/06/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	09/06/2023	09/06/2023
					CT2985	Active	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	09/06/2023	09/06/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	09/06/2023	09/06/2023
					CT3001	Implantable metal	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	09/06/2023	09/06/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	09/06/2023	09/06/2023
					CT3002	Implantable synthetic polymer	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	09/06/2023	09/06/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	09/06/2023	09/06/2023
					CT3006	Implantable inorganic material	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	09/06/2023	09/06/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	09/06/2023	09/06/2023
					CT314	Electrically powered	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	09/06/2023	09/06/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	09/06/2023	09/06/2023
					CT321	Long-term surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	09/06/2023	09/06/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	09/06/2023	09/06/2023
					CT449	Generators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	09/06/2023	09/06/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	09/06/2023	09/06/2023
					CT981	Single use	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	09/06/2023	09/06/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	09/06/2023	09/06/2023
					CT983	Surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	09/06/2023	09/06/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	09/06/2023	09/06/2023
13978	2023	All other reportable incidents	47263	Cardiac resynchronization therapy implantable pacemaker	CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	13/06/2023	13/06/2023
							F1901	Additional Surgery	A070504	Premature Discharge of Battery	13/06/2023	13/06/2023
					CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	13/06/2023	13/06/2023
							F1901	Additional Surgery	A070504	Premature Discharge of Battery	13/06/2023	13/06/2023
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	13/06/2023	13/06/2023
							F1901	Additional Surgery	A070504	Premature Discharge of Battery	13/06/2023	13/06/2023
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	13/06/2023	13/06/2023
							F1901	Additional Surgery	A070504	Premature Discharge of Battery	13/06/2023	13/06/2023
13979	2023	All other reportable incidents	47270	Cardiac resynchronization therapy implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A070912	Incorrect Interpretation of Signal	22/05/2023	22/05/2023
									A071302	Inappropriate/Inadequate Shock/Stimulation	22/05/2023	22/05/2023
							F1901	Additional Surgery	A070912	Incorrect Interpretation of Signal	22/05/2023	22/05/2023
									A071302	Inappropriate/Inadequate Shock/Stimulation	22/05/2023	22/05/2023
13980	2023	All other reportable incidents	47265	Dual-chamber implantable pacemaker, rate-responsive	CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	11/08/2023	11/08/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	11/08/2023	11/08/2023
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	11/08/2023	11/08/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	11/08/2023	11/08/2023
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	11/08/2023	11/08/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	11/08/2023	11/08/2023
13981	2023	All other reportable incidents	35223	Endocardial/interventricular septal pacing lead	CT1684	Cardiac pacemakers and associated devices	E0604	Cardiac Perforation	A070102	High Capture Threshold	05/09/2023	05/09/2023
									A072201	High Impedance	05/09/2023	05/09/2023
									A24	No Apparent Device Problem	05/09/2023	05/09/2023
							F08	Hospitalization or Prolonged Hospitalization	A070102	High Capture Threshold	05/09/2023	05/09/2023
									A072201	High Impedance	05/09/2023	05/09/2023
									A24	No Apparent Device Problem	05/09/2023	05/09/2023
							F1203	Life Threatening Illness or Injury	A070102	High Capture Threshold	05/09/2023	05/09/2023
									A072201	High Impedance	05/09/2023	05/09/2023
									A24	No Apparent Device Problem	05/09/2023	05/09/2023
							F1905	Device Revision or Replacement	A070102	High Capture Threshold	05/09/2023	05/09/2023
									A072201	High Impedance	05/09/2023	05/09/2023
									A24	No Apparent Device Problem	05/09/2023	05/09/2023
							F23	Unexpected Medical Intervention	A070102	High Capture Threshold	05/09/2023	05/09/2023
									A072201	High Impedance	05/09/2023	05/09/2023
									A24	No Apparent Device Problem	05/09/2023	05/09/2023
13982	2023	All other reportable incidents	35852	Single-chamber implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery		
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery		
13983	2023	All other reportable incidents	35852	Single-chamber implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E2104	Electric Shock	A0712	Pacing Problem		
									A071302	Inappropriate/Inadequate Shock/Stimulation		
							F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A0712	Pacing Problem		
									A071302	Inappropriate/Inadequate Shock/Stimulation		
13984	2023	All other reportable incidents	37265	Dual-chamber implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A072201	High Impedance		
							F26	No Health Consequences or Impact	A072201	High Impedance		
13985	2023	All other reportable incidents	47265	Dual-chamber implantable pacemaker, rate-responsive	CT141	Electrodes/Leads and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A0908	Incorrect, Inadequate or Imprecise Result or Readings	09/06/2023	09/06/2023

							F08	Hospitalization or Prolonged Hospitalization	A0908	Incorrect, Inadequate or Imprecise Result or Readings	09/06/2023	09/06/2023
							F1905	Device Revision or Replacement	A0908	Incorrect, Inadequate or Imprecise Result or Readings	09/06/2023	09/06/2023
					CT145	Cardiovascular devices	E2403	No Clinical Signs, Symptoms or Conditions	A0908	Incorrect, Inadequate or Imprecise Result or Readings	09/06/2023	09/06/2023
							F08	Hospitalization or Prolonged Hospitalization	A0908	Incorrect, Inadequate or Imprecise Result or Readings	09/06/2023	09/06/2023
							F1905	Device Revision or Replacement	A0908	Incorrect, Inadequate or Imprecise Result or Readings	09/06/2023	09/06/2023
					CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A0908	Incorrect, Inadequate or Imprecise Result or Readings	09/06/2023	09/06/2023
							F08	Hospitalization or Prolonged Hospitalization	A0908	Incorrect, Inadequate or Imprecise Result or Readings	09/06/2023	09/06/2023
							F1905	Device Revision or Replacement	A0908	Incorrect, Inadequate or Imprecise Result or Readings	09/06/2023	09/06/2023
					CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A0908	Incorrect, Inadequate or Imprecise Result or Readings	09/06/2023	09/06/2023
							F08	Hospitalization or Prolonged Hospitalization	A0908	Incorrect, Inadequate or Imprecise Result or Readings	09/06/2023	09/06/2023
							F1905	Device Revision or Replacement	A0908	Incorrect, Inadequate or Imprecise Result or Readings	09/06/2023	09/06/2023
					CT233	Surgical	E2403	No Clinical Signs, Symptoms or Conditions	A0908	Incorrect, Inadequate or Imprecise Result or Readings	09/06/2023	09/06/2023
							F08	Hospitalization or Prolonged Hospitalization	A0908	Incorrect, Inadequate or Imprecise Result or Readings	09/06/2023	09/06/2023
							F1905	Device Revision or Replacement	A0908	Incorrect, Inadequate or Imprecise Result or Readings	09/06/2023	09/06/2023
					CT2406	Implantable	E2403	No Clinical Signs, Symptoms or Conditions	A0908	Incorrect, Inadequate or Imprecise Result or Readings	09/06/2023	09/06/2023
							F08	Hospitalization or Prolonged Hospitalization	A0908	Incorrect, Inadequate or Imprecise Result or Readings	09/06/2023	09/06/2023
							F1905	Device Revision or Replacement	A0908	Incorrect, Inadequate or Imprecise Result or Readings	09/06/2023	09/06/2023
					CT2407	Active implantable	E2403	No Clinical Signs, Symptoms or Conditions	A0908	Incorrect, Inadequate or Imprecise Result or Readings	09/06/2023	09/06/2023
							F08	Hospitalization or Prolonged Hospitalization	A0908	Incorrect, Inadequate or Imprecise Result or Readings	09/06/2023	09/06/2023
							F1905	Device Revision or Replacement	A0908	Incorrect, Inadequate or Imprecise Result or Readings	09/06/2023	09/06/2023
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A0908	Incorrect, Inadequate or Imprecise Result or Readings	09/06/2023	09/06/2023
							F08	Hospitalization or Prolonged Hospitalization	A0908	Incorrect, Inadequate or Imprecise Result or Readings	09/06/2023	09/06/2023
							F1905	Device Revision or Replacement	A0908	Incorrect, Inadequate or Imprecise Result or Readings	09/06/2023	09/06/2023
					CT2727	Active implant devices/ancillaries	E2403	No Clinical Signs, Symptoms or Conditions	A0908	Incorrect, Inadequate or Imprecise Result or Readings	09/06/2023	09/06/2023
							F08	Hospitalization or Prolonged Hospitalization	A0908	Incorrect, Inadequate or Imprecise Result or Readings	09/06/2023	09/06/2023
							F1905	Device Revision or Replacement	A0908	Incorrect, Inadequate or Imprecise Result or Readings	09/06/2023	09/06/2023
					CT2750	Implantable cardiac electrical stimulators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A0908	Incorrect, Inadequate or Imprecise Result or Readings	09/06/2023	09/06/2023
							F08	Hospitalization or Prolonged Hospitalization	A0908	Incorrect, Inadequate or Imprecise Result or Readings	09/06/2023	09/06/2023
							F1905	Device Revision or Replacement	A0908	Incorrect, Inadequate or Imprecise Result or Readings	09/06/2023	09/06/2023
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A0908	Incorrect, Inadequate or Imprecise Result or Readings	09/06/2023	09/06/2023
							F08	Hospitalization or Prolonged Hospitalization	A0908	Incorrect, Inadequate or Imprecise Result or Readings	09/06/2023	09/06/2023
							F1905	Device Revision or Replacement	A0908	Incorrect, Inadequate or Imprecise Result or Readings	09/06/2023	09/06/2023
					CT2985	Active	E2403	No Clinical Signs, Symptoms or Conditions	A0908	Incorrect, Inadequate or Imprecise Result or Readings	09/06/2023	09/06/2023
							F08	Hospitalization or Prolonged Hospitalization	A0908	Incorrect, Inadequate or Imprecise Result or Readings	09/06/2023	09/06/2023
							F1905	Device Revision or Replacement	A0908	Incorrect, Inadequate or Imprecise Result or Readings	09/06/2023	09/06/2023
					CT3001	Implantable metal	E2403	No Clinical Signs, Symptoms or Conditions	A0908	Incorrect, Inadequate or Imprecise Result or Readings	09/06/2023	09/06/2023
							F08	Hospitalization or Prolonged Hospitalization	A0908	Incorrect, Inadequate or Imprecise Result or Readings	09/06/2023	09/06/2023
							F1905	Device Revision or Replacement	A0908	Incorrect, Inadequate or Imprecise Result or Readings	09/06/2023	09/06/2023
					CT3002	Implantable synthetic polymer	E2403	No Clinical Signs, Symptoms or Conditions	A0908	Incorrect, Inadequate or Imprecise Result or Readings	09/06/2023	09/06/2023
							F08	Hospitalization or Prolonged Hospitalization	A0908	Incorrect, Inadequate or Imprecise Result or Readings	09/06/2023	09/06/2023
							F1905	Device Revision or Replacement	A0908	Incorrect, Inadequate or Imprecise Result or Readings	09/06/2023	09/06/2023
					CT3006	Implantable inorganic material	E2403	No Clinical Signs, Symptoms or Conditions	A0908	Incorrect, Inadequate or Imprecise Result or Readings	09/06/2023	09/06/2023
							F08	Hospitalization or Prolonged Hospitalization	A0908	Incorrect, Inadequate or Imprecise Result or Readings	09/06/2023	09/06/2023
							F1905	Device Revision or Replacement	A0908	Incorrect, Inadequate or Imprecise Result or Readings	09/06/2023	09/06/2023
					CT314	Electrically powered	E2403	No Clinical Signs, Symptoms or Conditions	A0908	Incorrect, Inadequate or Imprecise Result or Readings	09/06/2023	09/06/2023
							F08	Hospitalization or Prolonged Hospitalization	A0908	Incorrect, Inadequate or Imprecise Result or Readings	09/06/2023	09/06/2023
							F1905	Device Revision or Replacement	A0908	Incorrect, Inadequate or Imprecise Result or Readings	09/06/2023	09/06/2023
					CT321	Long-term surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A0908	Incorrect, Inadequate or Imprecise Result or Readings	09/06/2023	09/06/2023
							F08	Hospitalization or Prolonged Hospitalization	A0908	Incorrect, Inadequate or Imprecise Result or Readings	09/06/2023	09/06/2023
							F1905	Device Revision or Replacement	A0908	Incorrect, Inadequate or Imprecise Result or Readings	09/06/2023	09/06/2023

					CT449	Generators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A0908	Incorrect, Inadequate or Imprecise Result or Readings	09/06/2023	09/06/2023
							F08	Hospitalization or Prolonged Hospitalization	A0908	Incorrect, Inadequate or Imprecise Result or Readings	09/06/2023	09/06/2023
							F1905	Device Revision or Replacement	A0908	Incorrect, Inadequate or Imprecise Result or Readings	09/06/2023	09/06/2023
					CT981	Single use	E2403	No Clinical Signs, Symptoms or Conditions	A0908	Incorrect, Inadequate or Imprecise Result or Readings	09/06/2023	09/06/2023
							F08	Hospitalization or Prolonged Hospitalization	A0908	Incorrect, Inadequate or Imprecise Result or Readings	09/06/2023	09/06/2023
							F1905	Device Revision or Replacement	A0908	Incorrect, Inadequate or Imprecise Result or Readings	09/06/2023	09/06/2023
					CT983	Surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A0908	Incorrect, Inadequate or Imprecise Result or Readings	09/06/2023	09/06/2023
							F08	Hospitalization or Prolonged Hospitalization	A0908	Incorrect, Inadequate or Imprecise Result or Readings	09/06/2023	09/06/2023
							F1905	Device Revision or Replacement	A0908	Incorrect, Inadequate or Imprecise Result or Readings	09/06/2023	09/06/2023
13986	2023	All other reportable incidents	17882	Physiologic-monitoring defibrillation system	CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A071204	No Pacing		
							F27	Problem Identified During Non-Clinical Procedure	A071204	No Pacing		
13987	2023	All other reportable incidents	35852	Single-chamber implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
									A0908	Incorrect, Inadequate or Imprecise Result or Readings		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
									A0908	Incorrect, Inadequate or Imprecise Result or Readings		
13988	2023	All other reportable incidents	35852	Single-chamber implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E2104	Electric Shock	A070909	Oversensing		
									A071302	Inappropriate/Inadequate Shock/Stimulation		
									A1301	Failure to Read Input Signal		
							F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A070909	Oversensing		
									A071302	Inappropriate/Inadequate Shock/Stimulation		
									A1301	Failure to Read Input Signal		
13989	2023	All other reportable incidents	17882	Physiologic-monitoring defibrillation system	CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A071204	No Pacing		
							F27	Problem Identified During Non-Clinical Procedure	A071204	No Pacing		
13990	2023	All other reportable incidents	17882	Physiologic-monitoring defibrillation system	CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A071204	No Pacing		
							F27	Problem Identified During Non-Clinical Procedure	A071204	No Pacing		
13991	2023		47265	Dual-chamber implantable pacemaker, rate-responsive	CT141	Electrodes/Leads and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	07/08/2023	07/08/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	07/08/2023	07/08/2023
					CT145	Cardiovascular devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	07/08/2023	07/08/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	07/08/2023	07/08/2023
					CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	07/08/2023	07/08/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	07/08/2023	07/08/2023
					CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	07/08/2023	07/08/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	07/08/2023	07/08/2023
					CT233	Surgical	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	07/08/2023	07/08/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	07/08/2023	07/08/2023
					CT2406	Implantable	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	07/08/2023	07/08/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	07/08/2023	07/08/2023
					CT2407	Active implantable	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	07/08/2023	07/08/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	07/08/2023	07/08/2023
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	07/08/2023	07/08/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	07/08/2023	07/08/2023
					CT2727	Active implant devices/ancillaries	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	07/08/2023	07/08/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	07/08/2023	07/08/2023
					CT2750	Implantable cardiac electrical stimulators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	07/08/2023	07/08/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	07/08/2023	07/08/2023
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	07/08/2023	07/08/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	07/08/2023	07/08/2023
					CT2985	Active	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	07/08/2023	07/08/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	07/08/2023	07/08/2023
					CT3001	Implantable metal	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	07/08/2023	07/08/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	07/08/2023	07/08/2023
					CT3002	Implantable synthetic polymer	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	07/08/2023	07/08/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	07/08/2023	07/08/2023
					CT3006	Implantable inorganic material	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	07/08/2023	07/08/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	07/08/2023	07/08/2023
					CT314	Electrically powered	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	07/08/2023	07/08/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	07/08/2023	07/08/2023
					CT321	Long-term surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	07/08/2023	07/08/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	07/08/2023	07/08/2023
					CT449	Generators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	07/08/2023	07/08/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	07/08/2023	07/08/2023
					CT981	Single use	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	07/08/2023	07/08/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	07/08/2023	07/08/2023
					CT983	Surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	07/08/2023	07/08/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	07/08/2023	07/08/2023
13992	2023		47265	Dual-chamber implantable pacemaker, rate-responsive	CT141	Electrodes/Leads and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	07/08/2023	07/08/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	07/08/2023	07/08/2023
					CT145	Cardiovascular devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	07/08/2023	07/08/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	07/08/2023	07/08/2023
					CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	07/08/2023	07/08/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	07/08/2023	07/08/2023
					CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	07/08/2023	07/08/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	07/08/2023	07/08/2023
					CT233	Surgical	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	07/08/2023	07/08/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	07/08/2023	07/08/2023
					CT2406	Implantable	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	07/08/2023	07/08/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	07/08/2023	07/08/2023
					CT2407	Active implantable	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	07/08/2023	07/08/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	07/08/2023	07/08/2023

					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	07/08/2023	07/08/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	07/08/2023	07/08/2023
					CT2727	Active implant devices/ancillaries	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	07/08/2023	07/08/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	07/08/2023	07/08/2023
					CT2750	Implantable cardiac electrical stimulators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	07/08/2023	07/08/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	07/08/2023	07/08/2023
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	07/08/2023	07/08/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	07/08/2023	07/08/2023
					CT2985	Active	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	07/08/2023	07/08/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	07/08/2023	07/08/2023
					CT3001	Implantable metal	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	07/08/2023	07/08/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	07/08/2023	07/08/2023
					CT3002	Implantable synthetic polymer	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	07/08/2023	07/08/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	07/08/2023	07/08/2023
					CT3006	Implantable inorganic material	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	07/08/2023	07/08/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	07/08/2023	07/08/2023
					CT314	Electrically powered	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	07/08/2023	07/08/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	07/08/2023	07/08/2023
					CT321	Long-term surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	07/08/2023	07/08/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	07/08/2023	07/08/2023
					CT449	Generators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	07/08/2023	07/08/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	07/08/2023	07/08/2023
					CT981	Single use	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	07/08/2023	07/08/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	07/08/2023	07/08/2023
					CT983	Surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	07/08/2023	07/08/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	07/08/2023	07/08/2023
13993	2023	All other reportable incidents	35223	Endocardial/interventricular septal pacing lead	CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070909	Oversensing		
							F26	No Health Consequences or Impact	A070909	Oversensing		
13994	2023		47265	Dual-chamber implantable pacemaker, rate-responsive	CT141	Electrodes/Leads and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	20/06/2023	20/06/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	20/06/2023	20/06/2023
					CT145	Cardiovascular devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	20/06/2023	20/06/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	20/06/2023	20/06/2023
					CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	20/06/2023	20/06/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	20/06/2023	20/06/2023
					CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	20/06/2023	20/06/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	20/06/2023	20/06/2023
					CT233	Surgical	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	20/06/2023	20/06/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	20/06/2023	20/06/2023
					CT2406	Implantable	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	20/06/2023	20/06/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	20/06/2023	20/06/2023
					CT2407	Active implantable	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	20/06/2023	20/06/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	20/06/2023	20/06/2023
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	20/06/2023	20/06/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	20/06/2023	20/06/2023
					CT2727	Active implant devices/ancillaries	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	20/06/2023	20/06/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	20/06/2023	20/06/2023
					CT2750	Implantable cardiac electrical stimulators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	20/06/2023	20/06/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	20/06/2023	20/06/2023
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	20/06/2023	20/06/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	20/06/2023	20/06/2023
					CT2985	Active	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	20/06/2023	20/06/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	20/06/2023	20/06/2023
					CT3001	Implantable metal	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	20/06/2023	20/06/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	20/06/2023	20/06/2023
					CT3002	Implantable synthetic polymer	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	20/06/2023	20/06/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	20/06/2023	20/06/2023
					CT3006	Implantable inorganic material	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	20/06/2023	20/06/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	20/06/2023	20/06/2023
					CT314	Electrically powered	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	20/06/2023	20/06/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	20/06/2023	20/06/2023
					CT321	Long-term surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	20/06/2023	20/06/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	20/06/2023	20/06/2023
					CT449	Generators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	20/06/2023	20/06/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	20/06/2023	20/06/2023
					CT981	Single use	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	20/06/2023	20/06/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	20/06/2023	20/06/2023
					CT983	Surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	20/06/2023	20/06/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	20/06/2023	20/06/2023
13995	2023	All other reportable incidents	17882	Physiologic-monitoring defibrillation system	CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070803	Failure to Power Up		
							F27	Problem Identified During Non-Clinical Procedure	A070803	Failure to Power Up		
13996	2023	All other reportable incidents	47265	Dual-chamber implantable pacemaker, rate-responsive	CT141	Electrodes/Leads and associated devices	E0112	Dizziness	A070504	Premature Discharge of Battery	22/09/2023	22/09/2023
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	22/09/2023	22/09/2023
							F12	Serious Injury/ Illness/ Impairment	A070504	Premature Discharge of Battery	22/09/2023	22/09/2023
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	22/09/2023	22/09/2023
					CT145	Cardiovascular devices	E0112	Dizziness	A070504	Premature Discharge of Battery	22/09/2023	22/09/2023
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	22/09/2023	22/09/2023
							F12	Serious Injury/ Illness/ Impairment	A070504	Premature Discharge of Battery	22/09/2023	22/09/2023
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	22/09/2023	22/09/2023
					CT1684	Cardiac pacemakers and associated devices	E0112	Dizziness	A070504	Premature Discharge of Battery	22/09/2023	22/09/2023
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	22/09/2023	22/09/2023
							F12	Serious Injury/ Illness/ Impairment	A070504	Premature Discharge of Battery	22/09/2023	22/09/2023
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	22/09/2023	22/09/2023
					CT1685	Cardiac pacemakers	E0112	Dizziness	A070504	Premature Discharge of Battery	22/09/2023	22/09/2023
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	22/09/2023	22/09/2023
							F12	Serious Injury/ Illness/ Impairment	A070504	Premature Discharge of Battery	22/09/2023	22/09/2023
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	22/09/2023	22/09/2023
					CT233	Surgical	E0112	Dizziness	A070504	Premature Discharge of Battery	22/09/2023	22/09/2023
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	22/09/2023	22/09/2023

							F12	Serious Injury/ Illness/ Impairment	A070504	Premature Discharge of Battery	22/09/2023	22/09/2023
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	22/09/2023	22/09/2023
					CT2406	Implantable	E0112	Dizziness	A070504	Premature Discharge of Battery	22/09/2023	22/09/2023
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	22/09/2023	22/09/2023
							F12	Serious Injury/ Illness/ Impairment	A070504	Premature Discharge of Battery	22/09/2023	22/09/2023
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	22/09/2023	22/09/2023
					CT2407	Active implantable	E0112	Dizziness	A070504	Premature Discharge of Battery	22/09/2023	22/09/2023
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	22/09/2023	22/09/2023
							F12	Serious Injury/ Illness/ Impairment	A070504	Premature Discharge of Battery	22/09/2023	22/09/2023
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	22/09/2023	22/09/2023
					CT257	Implantable cardiac pacemakers	E0112	Dizziness	A070504	Premature Discharge of Battery	22/09/2023	22/09/2023
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	22/09/2023	22/09/2023
							F12	Serious Injury/ Illness/ Impairment	A070504	Premature Discharge of Battery	22/09/2023	22/09/2023
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	22/09/2023	22/09/2023
					CT2727	Active implant devices/ancillaries	E0112	Dizziness	A070504	Premature Discharge of Battery	22/09/2023	22/09/2023
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	22/09/2023	22/09/2023
							F12	Serious Injury/ Illness/ Impairment	A070504	Premature Discharge of Battery	22/09/2023	22/09/2023
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	22/09/2023	22/09/2023
					CT2750	Implantable cardiac electrical stimulators and associated devices	E0112	Dizziness	A070504	Premature Discharge of Battery	22/09/2023	22/09/2023
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	22/09/2023	22/09/2023
							F12	Serious Injury/ Illness/ Impairment	A070504	Premature Discharge of Battery	22/09/2023	22/09/2023
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	22/09/2023	22/09/2023
					CT2866	Implantable cardiac electrical stimulators	E0112	Dizziness	A070504	Premature Discharge of Battery	22/09/2023	22/09/2023
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	22/09/2023	22/09/2023
							F12	Serious Injury/ Illness/ Impairment	A070504	Premature Discharge of Battery	22/09/2023	22/09/2023
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	22/09/2023	22/09/2023
					CT2985	Active	E0112	Dizziness	A070504	Premature Discharge of Battery	22/09/2023	22/09/2023
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	22/09/2023	22/09/2023
							F12	Serious Injury/ Illness/ Impairment	A070504	Premature Discharge of Battery	22/09/2023	22/09/2023
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	22/09/2023	22/09/2023
					CT3001	Implantable metal	E0112	Dizziness	A070504	Premature Discharge of Battery	22/09/2023	22/09/2023
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	22/09/2023	22/09/2023
							F12	Serious Injury/ Illness/ Impairment	A070504	Premature Discharge of Battery	22/09/2023	22/09/2023
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	22/09/2023	22/09/2023
					CT3002	Implantable synthetic polymer	E0112	Dizziness	A070504	Premature Discharge of Battery	22/09/2023	22/09/2023
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	22/09/2023	22/09/2023
							F12	Serious Injury/ Illness/ Impairment	A070504	Premature Discharge of Battery	22/09/2023	22/09/2023
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	22/09/2023	22/09/2023
					CT3006	Implantable inorganic material	E0112	Dizziness	A070504	Premature Discharge of Battery	22/09/2023	22/09/2023
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	22/09/2023	22/09/2023
							F12	Serious Injury/ Illness/ Impairment	A070504	Premature Discharge of Battery	22/09/2023	22/09/2023
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	22/09/2023	22/09/2023
					CT314	Electrically powered	E0112	Dizziness	A070504	Premature Discharge of Battery	22/09/2023	22/09/2023
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	22/09/2023	22/09/2023
							F12	Serious Injury/ Illness/ Impairment	A070504	Premature Discharge of Battery	22/09/2023	22/09/2023
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	22/09/2023	22/09/2023
					CT321	Long-term surgical invasive	E0112	Dizziness	A070504	Premature Discharge of Battery	22/09/2023	22/09/2023
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	22/09/2023	22/09/2023
							F12	Serious Injury/ Illness/ Impairment	A070504	Premature Discharge of Battery	22/09/2023	22/09/2023
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	22/09/2023	22/09/2023
					CT449	Generators and associated devices	E0112	Dizziness	A070504	Premature Discharge of Battery	22/09/2023	22/09/2023
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	22/09/2023	22/09/2023
							F12	Serious Injury/ Illness/ Impairment	A070504	Premature Discharge of Battery	22/09/2023	22/09/2023
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	22/09/2023	22/09/2023
					CT981	Single use	E0112	Dizziness	A070504	Premature Discharge of Battery	22/09/2023	22/09/2023
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	22/09/2023	22/09/2023
							F12	Serious Injury/ Illness/ Impairment	A070504	Premature Discharge of Battery	22/09/2023	22/09/2023
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	22/09/2023	22/09/2023
					CT983	Surgical invasive	E0112	Dizziness	A070504	Premature Discharge of Battery	22/09/2023	22/09/2023
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	22/09/2023	22/09/2023
							F12	Serious Injury/ Illness/ Impairment	A070504	Premature Discharge of Battery	22/09/2023	22/09/2023
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	22/09/2023	22/09/2023
13997	2023	All other reportable incidents	35852	Single-chamber implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E2104	Electric Shock	A070909	Oversensing	01/12/2022	01/12/2022
									A090801	Signal Artifact/Noise	01/12/2022	01/12/2022
									A1301	Failure to Read Input Signal	01/12/2022	01/12/2022
							F08	Hospitalization or Prolonged Hospitalization	A070909	Oversensing	01/12/2022	01/12/2022
									A090801	Signal Artifact/Noise	01/12/2022	01/12/2022
									A1301	Failure to Read Input Signal	01/12/2022	01/12/2022
							F1905	Device Revision or Replacement	A070909	Oversensing	01/12/2022	01/12/2022
									A090801	Signal Artifact/Noise	01/12/2022	01/12/2022
									A1301	Failure to Read Input Signal	01/12/2022	01/12/2022
							F2203	Imaging Required	A070909	Oversensing	01/12/2022	01/12/2022
									A090801	Signal Artifact/Noise	01/12/2022	01/12/2022
									A1301	Failure to Read Input Signal	01/12/2022	01/12/2022
13998	2023	All other reportable incidents	47265	Dual-chamber implantable pacemaker, rate-responsive	CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A0908	Incorrect, Inadequate or Imprecise Result or Readings	22/08/2023	22/08/2023
							F08	Hospitalization or Prolonged Hospitalization	A0908	Incorrect, Inadequate or Imprecise Result or Readings	22/08/2023	22/08/2023
							F19	Surgical Intervention	A0908	Incorrect, Inadequate or Imprecise Result or Readings	22/08/2023	22/08/2023
							F1903	Device Explantation	A0908	Incorrect, Inadequate or Imprecise Result or Readings	22/08/2023	22/08/2023
							F1905	Device Revision or Replacement	A0908	Incorrect, Inadequate or Imprecise Result or Readings	22/08/2023	22/08/2023
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A0908	Incorrect, Inadequate or Imprecise Result or Readings	22/08/2023	22/08/2023
							F08	Hospitalization or Prolonged Hospitalization	A0908	Incorrect, Inadequate or Imprecise Result or Readings	22/08/2023	22/08/2023
							F19	Surgical Intervention	A0908	Incorrect, Inadequate or Imprecise Result or Readings	22/08/2023	22/08/2023

							F1903	Device Explantation	A0908	Incorrect, Inadequate or Imprecise Result or Readings	22/08/2023	22/08/2023
							F1905	Device Revision or Replacement	A0908	Incorrect, Inadequate or Imprecise Result or Readings	22/08/2023	22/08/2023
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A0908	Incorrect, Inadequate or Imprecise Result or Readings	22/08/2023	22/08/2023
							F08	Hospitalization or Prolonged Hospitalization	A0908	Incorrect, Inadequate or Imprecise Result or Readings	22/08/2023	22/08/2023
							F19	Surgical Intervention	A0908	Incorrect, Inadequate or Imprecise Result or Readings	22/08/2023	22/08/2023
							F1903	Device Explantation	A0908	Incorrect, Inadequate or Imprecise Result or Readings	22/08/2023	22/08/2023
							F1905	Device Revision or Replacement	A0908	Incorrect, Inadequate or Imprecise Result or Readings	22/08/2023	22/08/2023
13999	2023		47265	Dual-chamber implantable pacemaker, rate-responsive	CT141	Electrodes/Leads and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem		
							F1901	Additional Surgery	A24	No Apparent Device Problem		
					CT145	Cardiovascular devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem		
							F1901	Additional Surgery	A24	No Apparent Device Problem		
					CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem		
							F1901	Additional Surgery	A24	No Apparent Device Problem		
					CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem		
							F1901	Additional Surgery	A24	No Apparent Device Problem		
					CT233	Surgical	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem		
							F1901	Additional Surgery	A24	No Apparent Device Problem		
					CT2406	Implantable	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem		
							F1901	Additional Surgery	A24	No Apparent Device Problem		
					CT2407	Active implantable	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem		
							F1901	Additional Surgery	A24	No Apparent Device Problem		
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem		
							F1901	Additional Surgery	A24	No Apparent Device Problem		
					CT2727	Active implant devices/ancillaries	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem		
							F1901	Additional Surgery	A24	No Apparent Device Problem		
					CT2750	Implantable cardiac electrical stimulators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem		
							F1901	Additional Surgery	A24	No Apparent Device Problem		
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem		
							F1901	Additional Surgery	A24	No Apparent Device Problem		
					CT2985	Active	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem		
							F1901	Additional Surgery	A24	No Apparent Device Problem		
					CT3001	Implantable metal	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem		
							F1901	Additional Surgery	A24	No Apparent Device Problem		
					CT3002	Implantable synthetic polymer	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem		
							F1901	Additional Surgery	A24	No Apparent Device Problem		
					CT3006	Implantable inorganic material	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem		
							F1901	Additional Surgery	A24	No Apparent Device Problem		
					CT314	Electrically powered	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem		
							F1901	Additional Surgery	A24	No Apparent Device Problem		
					CT321	Long-term surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem		
							F1901	Additional Surgery	A24	No Apparent Device Problem		
					CT449	Generators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem		
							F1901	Additional Surgery	A24	No Apparent Device Problem		
					CT981	Single use	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem		
							F1901	Additional Surgery	A24	No Apparent Device Problem		
					CT983	Surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem		
							F1901	Additional Surgery	A24	No Apparent Device Problem		
14000	2023		47265	Dual-chamber implantable pacemaker, rate-responsive	CT141	Electrodes/Leads and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem		
							F1901	Additional Surgery	A24	No Apparent Device Problem		
					CT145	Cardiovascular devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem		
							F1901	Additional Surgery	A24	No Apparent Device Problem		
					CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem		
							F1901	Additional Surgery	A24	No Apparent Device Problem		
					CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem		
							F1901	Additional Surgery	A24	No Apparent Device Problem		
					CT233	Surgical	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem		
							F1901	Additional Surgery	A24	No Apparent Device Problem		
					CT2406	Implantable	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem		
							F1901	Additional Surgery	A24	No Apparent Device Problem		
					CT2407	Active implantable	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem		
							F1901	Additional Surgery	A24	No Apparent Device Problem		
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem		
							F1901	Additional Surgery	A24	No Apparent Device Problem		
					CT2727	Active implant devices/ancillaries	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem		
							F1901	Additional Surgery	A24	No Apparent Device Problem		
					CT2750	Implantable cardiac electrical stimulators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem		
							F1901	Additional Surgery	A24	No Apparent Device Problem		
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem		
							F1901	Additional Surgery	A24	No Apparent Device Problem		
					CT2985	Active	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem		
							F1901	Additional Surgery	A24	No Apparent Device Problem		
					CT3001	Implantable metal	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem		
							F1901	Additional Surgery	A24	No Apparent Device Problem		
					CT3002	Implantable synthetic polymer	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem		
							F1901	Additional Surgery	A24	No Apparent Device Problem		
					CT3006	Implantable inorganic material	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem		
							F1901	Additional Surgery	A24	No Apparent Device Problem		
					CT314	Electrically powered	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem		
							F1901	Additional Surgery	A24	No Apparent Device Problem		
					CT321	Long-term surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem		
							F1901	Additional Surgery	A24	No Apparent Device Problem		
					CT449	Generators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem		

						F1901	Additional Surgery	A24	No Apparent Device Problem		
				CT981	Single use	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem		
						F1901	Additional Surgery	A24	No Apparent Device Problem		
				CT983	Surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem		
						F1901	Additional Surgery	A24	No Apparent Device Problem		
14001	2023		47265	Dual-chamber implantable pacemaker, rate-responsive	CT141	Electrodes/Leads and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem		
							F1901	Additional Surgery	A24	No Apparent Device Problem		
					CT145	Cardiovascular devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem		
							F1901	Additional Surgery	A24	No Apparent Device Problem		
					CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem		
							F1901	Additional Surgery	A24	No Apparent Device Problem		
					CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem		
							F1901	Additional Surgery	A24	No Apparent Device Problem		
					CT233	Surgical	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem		
							F1901	Additional Surgery	A24	No Apparent Device Problem		
					CT2406	Implantable	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem		
							F1901	Additional Surgery	A24	No Apparent Device Problem		
					CT2407	Active implantable	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem		
							F1901	Additional Surgery	A24	No Apparent Device Problem		
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem		
							F1901	Additional Surgery	A24	No Apparent Device Problem		
					CT2727	Active implant devices/ancillaries	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem		
							F1901	Additional Surgery	A24	No Apparent Device Problem		
					CT2750	Implantable cardiac electrical stimulators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem		
							F1901	Additional Surgery	A24	No Apparent Device Problem		
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem		
							F1901	Additional Surgery	A24	No Apparent Device Problem		
					CT2985	Active	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem		
							F1901	Additional Surgery	A24	No Apparent Device Problem		
					CT3001	Implantable metal	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem		
							F1901	Additional Surgery	A24	No Apparent Device Problem		
					CT3002	Implantable synthetic polymer	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem		
							F1901	Additional Surgery	A24	No Apparent Device Problem		
					CT3006	Implantable inorganic material	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem		
							F1901	Additional Surgery	A24	No Apparent Device Problem		
					CT314	Electrically powered	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem		
							F1901	Additional Surgery	A24	No Apparent Device Problem		
					CT321	Long-term surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem		
							F1901	Additional Surgery	A24	No Apparent Device Problem		
					CT449	Generators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem		
							F1901	Additional Surgery	A24	No Apparent Device Problem		
					CT981	Single use	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem		
							F1901	Additional Surgery	A24	No Apparent Device Problem		
					CT983	Surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem		
							F1901	Additional Surgery	A24	No Apparent Device Problem		
14002	2023		47265	Dual-chamber implantable pacemaker, rate-responsive	CT141	Electrodes/Leads and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem		
							F1901	Additional Surgery	A24	No Apparent Device Problem		
					CT145	Cardiovascular devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem		
							F1901	Additional Surgery	A24	No Apparent Device Problem		
					CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem		
							F1901	Additional Surgery	A24	No Apparent Device Problem		
					CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem		
							F1901	Additional Surgery	A24	No Apparent Device Problem		
					CT233	Surgical	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem		
							F1901	Additional Surgery	A24	No Apparent Device Problem		
					CT2406	Implantable	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem		
							F1901	Additional Surgery	A24	No Apparent Device Problem		
					CT2407	Active implantable	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem		
							F1901	Additional Surgery	A24	No Apparent Device Problem		
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem		
							F1901	Additional Surgery	A24	No Apparent Device Problem		
					CT2727	Active implant devices/ancillaries	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem		
							F1901	Additional Surgery	A24	No Apparent Device Problem		
					CT2750	Implantable cardiac electrical stimulators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem		
							F1901	Additional Surgery	A24	No Apparent Device Problem		
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem		
							F1901	Additional Surgery	A24	No Apparent Device Problem		
					CT2985	Active	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem		
							F1901	Additional Surgery	A24	No Apparent Device Problem		
					CT3001	Implantable metal	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem		
							F1901	Additional Surgery	A24	No Apparent Device Problem		
					CT3002	Implantable synthetic polymer	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem		
							F1901	Additional Surgery	A24	No Apparent Device Problem		
					CT3006	Implantable inorganic material	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem		
							F1901	Additional Surgery	A24	No Apparent Device Problem		
					CT314	Electrically powered	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem		
							F1901	Additional Surgery	A24	No Apparent Device Problem		
					CT321	Long-term surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem		
							F1901	Additional Surgery	A24	No Apparent Device Problem		
					CT449	Generators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem		
							F1901	Additional Surgery	A24	No Apparent Device Problem		
					CT981	Single use	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem		
							F1901	Additional Surgery	A24	No Apparent Device Problem		
					CT983	Surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem		
							F1901	Additional Surgery	A24	No Apparent Device Problem		
14003	2023		47265	Dual-chamber implantable pacemaker, rate-responsive	CT141	Electrodes/Leads and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem		
							F1901	Additional Surgery	A24	No Apparent Device Problem		
					CT145	Cardiovascular devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem		

							F1901	Additional Surgery	A24	No Apparent Device Problem		
					CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem		
							F1901	Additional Surgery	A24	No Apparent Device Problem		
					CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem		
							F1901	Additional Surgery	A24	No Apparent Device Problem		
					CT233	Surgical	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem		
							F1901	Additional Surgery	A24	No Apparent Device Problem		
					CT2406	Implantable	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem		
							F1901	Additional Surgery	A24	No Apparent Device Problem		
					CT2407	Active implantable	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem		
							F1901	Additional Surgery	A24	No Apparent Device Problem		
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem		
							F1901	Additional Surgery	A24	No Apparent Device Problem		
					CT2727	Active implant devices/ancillaries	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem		
							F1901	Additional Surgery	A24	No Apparent Device Problem		
					CT2750	Implantable cardiac electrical stimulators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem		
							F1901	Additional Surgery	A24	No Apparent Device Problem		
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem		
							F1901	Additional Surgery	A24	No Apparent Device Problem		
					CT2985	Active	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem		
							F1901	Additional Surgery	A24	No Apparent Device Problem		
					CT3001	Implantable metal	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem		
							F1901	Additional Surgery	A24	No Apparent Device Problem		
					CT3002	Implantable synthetic polymer	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem		
							F1901	Additional Surgery	A24	No Apparent Device Problem		
					CT3006	Implantable inorganic material	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem		
							F1901	Additional Surgery	A24	No Apparent Device Problem		
					CT314	Electrically powered	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem		
							F1901	Additional Surgery	A24	No Apparent Device Problem		
					CT321	Long-term surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem		
							F1901	Additional Surgery	A24	No Apparent Device Problem		
					CT449	Generators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem		
							F1901	Additional Surgery	A24	No Apparent Device Problem		
					CT981	Single use	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem		
							F1901	Additional Surgery	A24	No Apparent Device Problem		
					CT983	Surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem		
							F1901	Additional Surgery	A24	No Apparent Device Problem		
14004	2023	All other reportable incidents	47265	Dual-chamber implantable pacemaker, rate-responsive	CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem		
							F1901	Additional Surgery	A24	No Apparent Device Problem		
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem		
							F1901	Additional Surgery	A24	No Apparent Device Problem		
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem		
							F1901	Additional Surgery	A24	No Apparent Device Problem		
14005	2023		47265	Dual-chamber implantable pacemaker, rate-responsive	CT141	Electrodes/Leads and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	16/05/2023	16/05/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	16/05/2023	16/05/2023
					CT145	Cardiovascular devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	16/05/2023	16/05/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	16/05/2023	16/05/2023
					CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	16/05/2023	16/05/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	16/05/2023	16/05/2023
					CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	16/05/2023	16/05/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	16/05/2023	16/05/2023
					CT233	Surgical	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	16/05/2023	16/05/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	16/05/2023	16/05/2023
					CT2406	Implantable	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	16/05/2023	16/05/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	16/05/2023	16/05/2023
					CT2407	Active implantable	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	16/05/2023	16/05/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	16/05/2023	16/05/2023
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	16/05/2023	16/05/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	16/05/2023	16/05/2023
					CT2727	Active implant devices/ancillaries	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	16/05/2023	16/05/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	16/05/2023	16/05/2023
					CT2750	Implantable cardiac electrical stimulators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	16/05/2023	16/05/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	16/05/2023	16/05/2023
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	16/05/2023	16/05/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	16/05/2023	16/05/2023
					CT2985	Active	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	16/05/2023	16/05/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	16/05/2023	16/05/2023
					CT3001	Implantable metal	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	16/05/2023	16/05/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	16/05/2023	16/05/2023
					CT3002	Implantable synthetic polymer	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	16/05/2023	16/05/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	16/05/2023	16/05/2023
					CT3006	Implantable inorganic material	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	16/05/2023	16/05/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	16/05/2023	16/05/2023
					CT314	Electrically powered	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	16/05/2023	16/05/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	16/05/2023	16/05/2023
					CT321	Long-term surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	16/05/2023	16/05/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	16/05/2023	16/05/2023
					CT449	Generators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	16/05/2023	16/05/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	16/05/2023	16/05/2023
					CT981	Single use	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	16/05/2023	16/05/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	16/05/2023	16/05/2023
					CT983	Surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	16/05/2023	16/05/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	16/05/2023	16/05/2023
14006	2023	All other reportable incidents	37265	Dual-chamber implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	31/07/2023	31/07/2023
							F1901	Additional Surgery	A070504	Premature Discharge of Battery	31/07/2023	31/07/2023
14007	2023	All other reportable incidents	17882	Physiologic-monitoring defibrillation system	CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem		
							F27	Problem Identified During Non-Clinical Procedure	A0705	Battery Problem		
14008	2023	All other reportable incidents	47265	Dual-chamber implantable pacemaker, rate-	CT141	Electrodes/Leads and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		

				responsive			F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT145	Cardiovascular devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT233	Surgical	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT2406	Implantable	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT2407	Active implantable	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT2727	Active implant devices/ancillaries	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT2750	Implantable cardiac electrical stimulators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT2985	Active	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT3001	Implantable metal	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT3002	Implantable synthetic polymer	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT3006	Implantable inorganic material	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT314	Electrically powered	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT321	Long-term surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT449	Generators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT981	Single use	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT983	Surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
14009	2023	All other reportable incidents	36241	Pacemaker/defibrillator lead	CT1684	Cardiac pacemakers and associated devices	E1906	Unspecified Infection	A24	No Apparent Device Problem		
							F08	Hospitalization or Prolonged Hospitalization	A24	No Apparent Device Problem		
							F1905	Device Revision or Replacement	A24	No Apparent Device Problem		
14010	2023	Death	47270	Cardiac resynchronization therapy implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E0601	Arrhythmia	A070910	Undersensing	16/10/2020	16/10/2020
									A070912	Incorrect Interpretation of Signal	16/10/2020	16/10/2020
									A071301	Failure to Deliver Shock/Stimulation	16/10/2020	16/10/2020
							F02	Death	A070910	Undersensing	16/10/2020	16/10/2020
									A070912	Incorrect Interpretation of Signal	16/10/2020	16/10/2020
									A071301	Failure to Deliver Shock/Stimulation	16/10/2020	16/10/2020
14011	2023	All other reportable incidents	60910	Coronary venous pacing lead	CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A0712	Pacing Problem		
							F12	Serious Injury/ Illness/ Impairment	A0712	Pacing Problem		
							F23	Unexpected Medical Intervention	A0712	Pacing Problem		
14012	2023	All other reportable incidents	17882	Physiologic-monitoring defibrillation system	CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A160503	Inappropriate or Unexpected Reset		
							F27	Problem Identified During Non-Clinical Procedure	A160503	Inappropriate or Unexpected Reset		
14013	2023	All other reportable incidents	47265	Dual-chamber implantable pacemaker, rate-responsive	CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	19/05/2023	19/05/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	19/05/2023	19/05/2023
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	19/05/2023	19/05/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	19/05/2023	19/05/2023
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	19/05/2023	19/05/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	19/05/2023	19/05/2023
14014	2023		17882	Physiologic-monitoring defibrillation system	CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A090204	Image Display Error/Artifact		
							F27	Problem Identified During Non-Clinical Procedure	A090204	Image Display Error/Artifact		
14015	2023	All other reportable incidents	36241	Pacemaker/defibrillator lead	CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070909	Oversensing		
									A090801	Signal Artifact/Noise		
							F26	No Health Consequences or Impact	A070909	Oversensing		
									A090801	Signal Artifact/Noise		
14016	2023	All other reportable incidents	36241	Pacemaker/defibrillator lead	CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070909	Oversensing		
									A090801	Signal Artifact/Noise		
							F26	No Health Consequences or Impact	A070909	Oversensing		
									A090801	Signal Artifact/Noise		
14017	2023	All other reportable incidents	37265	Dual-chamber implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E0601	Arrhythmia	A0712	Pacing Problem		
							F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A0712	Pacing Problem		
14018	2023	All other reportable incidents	47267	Single-chamber implantable pacemaker, rate-responsive	CT141	Electrodes/Leads and associated devices	E0601	Arrhythmia	A070910	Undersensing		
									A090801	Signal Artifact/Noise		
							E060102	Atrial Fibrillation	A070910	Undersensing		
									A090801	Signal Artifact/Noise		
							F26	No Health Consequences or Impact	A070910	Undersensing		
									A090801	Signal Artifact/Noise		
					CT145	Cardiovascular devices	E0601	Arrhythmia	A070910	Undersensing		
									A090801	Signal Artifact/Noise		
							E060102	Atrial Fibrillation	A070910	Undersensing		
									A090801	Signal Artifact/Noise		
							F26	No Health Consequences or Impact	A070910	Undersensing		
									A090801	Signal Artifact/Noise		
					CT1684	Cardiac pacemakers and associated devices	E0601	Arrhythmia	A070910	Undersensing		
									A090801	Signal Artifact/Noise		
							E060102	Atrial Fibrillation	A070910	Undersensing		
									A090801	Signal Artifact/Noise		
							F26	No Health Consequences or Impact	A070910	Undersensing		

									A090801	Signal Artifact/Noise		
					CT1685	Cardiac pacemakers	E0601	Arrhythmia	A070910	Undersensing		
									A090801	Signal Artifact/Noise		
							E060102	Atrial Fibrillation	A070910	Undersensing		
									A090801	Signal Artifact/Noise		
							F26	No Health Consequences or Impact	A070910	Undersensing		
									A090801	Signal Artifact/Noise		
					CT233	Surgical	E0601	Arrhythmia	A070910	Undersensing		
									A090801	Signal Artifact/Noise		
							E060102	Atrial Fibrillation	A070910	Undersensing		
									A090801	Signal Artifact/Noise		
							F26	No Health Consequences or Impact	A070910	Undersensing		
									A090801	Signal Artifact/Noise		
					CT2406	Implantable	E0601	Arrhythmia	A070910	Undersensing		
									A090801	Signal Artifact/Noise		
							E060102	Atrial Fibrillation	A070910	Undersensing		
									A090801	Signal Artifact/Noise		
							F26	No Health Consequences or Impact	A070910	Undersensing		
									A090801	Signal Artifact/Noise		
					CT2407	Active implantable	E0601	Arrhythmia	A070910	Undersensing		
									A090801	Signal Artifact/Noise		
							E060102	Atrial Fibrillation	A070910	Undersensing		
									A090801	Signal Artifact/Noise		
							F26	No Health Consequences or Impact	A070910	Undersensing		
									A090801	Signal Artifact/Noise		
					CT257	Implantable cardiac pacemakers	E0601	Arrhythmia	A070910	Undersensing		
									A090801	Signal Artifact/Noise		
							E060102	Atrial Fibrillation	A070910	Undersensing		
									A090801	Signal Artifact/Noise		
							F26	No Health Consequences or Impact	A070910	Undersensing		
									A090801	Signal Artifact/Noise		
					CT2727	Active implant devices/ancillaries	E0601	Arrhythmia	A070910	Undersensing		
									A090801	Signal Artifact/Noise		
							E060102	Atrial Fibrillation	A070910	Undersensing		
									A090801	Signal Artifact/Noise		
							F26	No Health Consequences or Impact	A070910	Undersensing		
									A090801	Signal Artifact/Noise		
					CT2750	Implantable cardiac electrical stimulators and associated devices	E0601	Arrhythmia	A070910	Undersensing		
									A090801	Signal Artifact/Noise		
							E060102	Atrial Fibrillation	A070910	Undersensing		
									A090801	Signal Artifact/Noise		
							F26	No Health Consequences or Impact	A070910	Undersensing		
									A090801	Signal Artifact/Noise		
					CT2866	Implantable cardiac electrical stimulators	E0601	Arrhythmia	A070910	Undersensing		
									A090801	Signal Artifact/Noise		
							E060102	Atrial Fibrillation	A070910	Undersensing		
									A090801	Signal Artifact/Noise		
							F26	No Health Consequences or Impact	A070910	Undersensing		
									A090801	Signal Artifact/Noise		
					CT2985	Active	E0601	Arrhythmia	A070910	Undersensing		
									A090801	Signal Artifact/Noise		
							E060102	Atrial Fibrillation	A070910	Undersensing		
									A090801	Signal Artifact/Noise		
							F26	No Health Consequences or Impact	A070910	Undersensing		
									A090801	Signal Artifact/Noise		
					CT3001	Implantable metal	E0601	Arrhythmia	A070910	Undersensing		
									A090801	Signal Artifact/Noise		
							E060102	Atrial Fibrillation	A070910	Undersensing		
									A090801	Signal Artifact/Noise		
							F26	No Health Consequences or Impact	A070910	Undersensing		
									A090801	Signal Artifact/Noise		
					CT3002	Implantable synthetic polymer	E0601	Arrhythmia	A070910	Undersensing		
									A090801	Signal Artifact/Noise		
							E060102	Atrial Fibrillation	A070910	Undersensing		
									A090801	Signal Artifact/Noise		
							F26	No Health Consequences or Impact	A070910	Undersensing		
									A090801	Signal Artifact/Noise		
					CT3006	Implantable inorganic material	E0601	Arrhythmia	A070910	Undersensing		
									A090801	Signal Artifact/Noise		
							E060102	Atrial Fibrillation	A070910	Undersensing		
									A090801	Signal Artifact/Noise		
							F26	No Health Consequences or Impact	A070910	Undersensing		
									A090801	Signal Artifact/Noise		
					CT314	Electrically powered	E0601	Arrhythmia	A070910	Undersensing		
									A090801	Signal Artifact/Noise		
							E060102	Atrial Fibrillation	A070910	Undersensing		
									A090801	Signal Artifact/Noise		
							F26	No Health Consequences or Impact	A070910	Undersensing		
									A090801	Signal Artifact/Noise		
					CT321	Long-term surgical invasive	E0601	Arrhythmia	A070910	Undersensing		
									A090801	Signal Artifact/Noise		
							E060102	Atrial Fibrillation	A070910	Undersensing		
									A090801	Signal Artifact/Noise		
							F26	No Health Consequences or Impact	A070910	Undersensing		
									A090801	Signal Artifact/Noise		
					CT449	Generators and associated devices	E0601	Arrhythmia	A070910	Undersensing		
									A090801	Signal Artifact/Noise		
							E060102	Atrial Fibrillation	A070910	Undersensing		

								A090801	Signal Artifact/Noise			
						F26	No Health Consequences or Impact	A070910	Undersensing			
								A090801	Signal Artifact/Noise			
					CT981	Single use	E0601	Arrhythmia	A070910	Undersensing		
									A090801	Signal Artifact/Noise		
							E060102	Atrial Fibrillation	A070910	Undersensing		
									A090801	Signal Artifact/Noise		
							F26	No Health Consequences or Impact	A070910	Undersensing		
									A090801	Signal Artifact/Noise		
					CT983	Surgical invasive	E0601	Arrhythmia	A070910	Undersensing		
									A090801	Signal Artifact/Noise		
							E060102	Atrial Fibrillation	A070910	Undersensing		
									A090801	Signal Artifact/Noise		
							F26	No Health Consequences or Impact	A070910	Undersensing		
									A090801	Signal Artifact/Noise		
14019	2023	All other reportable incidents	47265	Dual-chamber implantable pacemaker, rate-responsive	CT141	Electrodes/Leads and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A071102	Failure to Interrogate	06/04/2023	06/04/2023
							F1901	Additional Surgery	A071102	Failure to Interrogate	06/04/2023	06/04/2023
					CT145	Cardiovascular devices	E2403	No Clinical Signs, Symptoms or Conditions	A071102	Failure to Interrogate	06/04/2023	06/04/2023
							F1901	Additional Surgery	A071102	Failure to Interrogate	06/04/2023	06/04/2023
					CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A071102	Failure to Interrogate	06/04/2023	06/04/2023
							F1901	Additional Surgery	A071102	Failure to Interrogate	06/04/2023	06/04/2023
					CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A071102	Failure to Interrogate	06/04/2023	06/04/2023
							F1901	Additional Surgery	A071102	Failure to Interrogate	06/04/2023	06/04/2023
					CT233	Surgical	E2403	No Clinical Signs, Symptoms or Conditions	A071102	Failure to Interrogate	06/04/2023	06/04/2023
							F1901	Additional Surgery	A071102	Failure to Interrogate	06/04/2023	06/04/2023
					CT2406	Implantable	E2403	No Clinical Signs, Symptoms or Conditions	A071102	Failure to Interrogate	06/04/2023	06/04/2023
							F1901	Additional Surgery	A071102	Failure to Interrogate	06/04/2023	06/04/2023
					CT2407	Active implantable	E2403	No Clinical Signs, Symptoms or Conditions	A071102	Failure to Interrogate	06/04/2023	06/04/2023
							F1901	Additional Surgery	A071102	Failure to Interrogate	06/04/2023	06/04/2023
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A071102	Failure to Interrogate	06/04/2023	06/04/2023
							F1901	Additional Surgery	A071102	Failure to Interrogate	06/04/2023	06/04/2023
					CT2727	Active implant devices/ancillaries	E2403	No Clinical Signs, Symptoms or Conditions	A071102	Failure to Interrogate	06/04/2023	06/04/2023
							F1901	Additional Surgery	A071102	Failure to Interrogate	06/04/2023	06/04/2023
					CT2750	Implantable cardiac electrical stimulators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A071102	Failure to Interrogate	06/04/2023	06/04/2023
							F1901	Additional Surgery	A071102	Failure to Interrogate	06/04/2023	06/04/2023
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A071102	Failure to Interrogate	06/04/2023	06/04/2023
							F1901	Additional Surgery	A071102	Failure to Interrogate	06/04/2023	06/04/2023
					CT2985	Active	E2403	No Clinical Signs, Symptoms or Conditions	A071102	Failure to Interrogate	06/04/2023	06/04/2023
							F1901	Additional Surgery	A071102	Failure to Interrogate	06/04/2023	06/04/2023
					CT3001	Implantable metal	E2403	No Clinical Signs, Symptoms or Conditions	A071102	Failure to Interrogate	06/04/2023	06/04/2023
							F1901	Additional Surgery	A071102	Failure to Interrogate	06/04/2023	06/04/2023
					CT3002	Implantable synthetic polymer	E2403	No Clinical Signs, Symptoms or Conditions	A071102	Failure to Interrogate	06/04/2023	06/04/2023
							F1901	Additional Surgery	A071102	Failure to Interrogate	06/04/2023	06/04/2023
					CT3006	Implantable inorganic material	E2403	No Clinical Signs, Symptoms or Conditions	A071102	Failure to Interrogate	06/04/2023	06/04/2023
							F1901	Additional Surgery	A071102	Failure to Interrogate	06/04/2023	06/04/2023
					CT314	Electrically powered	E2403	No Clinical Signs, Symptoms or Conditions	A071102	Failure to Interrogate	06/04/2023	06/04/2023
							F1901	Additional Surgery	A071102	Failure to Interrogate	06/04/2023	06/04/2023
					CT321	Long-term surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A071102	Failure to Interrogate	06/04/2023	06/04/2023
							F1901	Additional Surgery	A071102	Failure to Interrogate	06/04/2023	06/04/2023
					CT449	Generators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A071102	Failure to Interrogate	06/04/2023	06/04/2023
							F1901	Additional Surgery	A071102	Failure to Interrogate	06/04/2023	06/04/2023
					CT981	Single use	E2403	No Clinical Signs, Symptoms or Conditions	A071102	Failure to Interrogate	06/04/2023	06/04/2023
							F1901	Additional Surgery	A071102	Failure to Interrogate	06/04/2023	06/04/2023
					CT983	Surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A071102	Failure to Interrogate	06/04/2023	06/04/2023
							F1901	Additional Surgery	A071102	Failure to Interrogate	06/04/2023	06/04/2023
14020	2023	All other reportable incidents	47265	Dual-chamber implantable pacemaker, rate-responsive	CT141	Electrodes/Leads and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070101	Failure to Capture		
									A070910	Undersensing		
							F26	No Health Consequences or Impact	A070101	Failure to Capture		
									A070910	Undersensing		
					CT145	Cardiovascular devices	E2403	No Clinical Signs, Symptoms or Conditions	A070101	Failure to Capture		
									A070910	Undersensing		
							F26	No Health Consequences or Impact	A070101	Failure to Capture		
									A070910	Undersensing		
					CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070101	Failure to Capture		
									A070910	Undersensing		
							F26	No Health Consequences or Impact	A070101	Failure to Capture		
									A070910	Undersensing		
					CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A070101	Failure to Capture		
									A070910	Undersensing		
							F26	No Health Consequences or Impact	A070101	Failure to Capture		
									A070910	Undersensing		
					CT233	Surgical	E2403	No Clinical Signs, Symptoms or Conditions	A070101	Failure to Capture		
									A070910	Undersensing		
							F26	No Health Consequences or Impact	A070101	Failure to Capture		
									A070910	Undersensing		
					CT2406	Implantable	E2403	No Clinical Signs, Symptoms or Conditions	A070101	Failure to Capture		
									A070910	Undersensing		
							F26	No Health Consequences or Impact	A070101	Failure to Capture		
									A070910	Undersensing		
					CT2407	Active implantable	E2403	No Clinical Signs, Symptoms or Conditions	A070101	Failure to Capture		
									A070910	Undersensing		
							F26	No Health Consequences or Impact	A070101	Failure to Capture		
									A070910	Undersensing		
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A070101	Failure to Capture		
									A070910	Undersensing		
							F26	No Health Consequences or Impact	A070101	Failure to Capture		
									A070910	Undersensing		
					CT2727	Active implant devices/ancillaries	E2403	No Clinical Signs, Symptoms or Conditions	A070101	Failure to Capture		

								A070910	Undersensing			
						F26	No Health Consequences or Impact	A070101	Failure to Capture			
								A070910	Undersensing			
					CT2750	Implantable cardiac electrical stimulators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070101	Failure to Capture		
									A070910	Undersensing		
							F26	No Health Consequences or Impact	A070101	Failure to Capture		
									A070910	Undersensing		
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A070101	Failure to Capture		
									A070910	Undersensing		
							F26	No Health Consequences or Impact	A070101	Failure to Capture		
									A070910	Undersensing		
					CT2985	Active	E2403	No Clinical Signs, Symptoms or Conditions	A070101	Failure to Capture		
									A070910	Undersensing		
							F26	No Health Consequences or Impact	A070101	Failure to Capture		
									A070910	Undersensing		
					CT3001	Implantable metal	E2403	No Clinical Signs, Symptoms or Conditions	A070101	Failure to Capture		
									A070910	Undersensing		
							F26	No Health Consequences or Impact	A070101	Failure to Capture		
									A070910	Undersensing		
					CT3002	Implantable synthetic polymer	E2403	No Clinical Signs, Symptoms or Conditions	A070101	Failure to Capture		
									A070910	Undersensing		
							F26	No Health Consequences or Impact	A070101	Failure to Capture		
									A070910	Undersensing		
					CT3006	Implantable inorganic material	E2403	No Clinical Signs, Symptoms or Conditions	A070101	Failure to Capture		
									A070910	Undersensing		
							F26	No Health Consequences or Impact	A070101	Failure to Capture		
									A070910	Undersensing		
					CT314	Electrically powered	E2403	No Clinical Signs, Symptoms or Conditions	A070101	Failure to Capture		
									A070910	Undersensing		
							F26	No Health Consequences or Impact	A070101	Failure to Capture		
									A070910	Undersensing		
					CT321	Long-term surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A070101	Failure to Capture		
									A070910	Undersensing		
							F26	No Health Consequences or Impact	A070101	Failure to Capture		
									A070910	Undersensing		
					CT449	Generators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070101	Failure to Capture		
									A070910	Undersensing		
							F26	No Health Consequences or Impact	A070101	Failure to Capture		
									A070910	Undersensing		
					CT981	Single use	E2403	No Clinical Signs, Symptoms or Conditions	A070101	Failure to Capture		
									A070910	Undersensing		
							F26	No Health Consequences or Impact	A070101	Failure to Capture		
									A070910	Undersensing		
					CT983	Surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A070101	Failure to Capture		
									A070910	Undersensing		
							F26	No Health Consequences or Impact	A070101	Failure to Capture		
									A070910	Undersensing		
14021	2023	All other reportable incidents	47263	Cardiac resynchronization therapy implantable pacemaker	CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A0908	Incorrect, Inadequate or Imprecise Result or Readings	01/09/2023	01/09/2023
							F08	Hospitalization or Prolonged Hospitalization	A0908	Incorrect, Inadequate or Imprecise Result or Readings	01/09/2023	01/09/2023
							F1905	Device Revision or Replacement	A0908	Incorrect, Inadequate or Imprecise Result or Readings	01/09/2023	01/09/2023
					CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A0908	Incorrect, Inadequate or Imprecise Result or Readings	01/09/2023	01/09/2023
							F08	Hospitalization or Prolonged Hospitalization	A0908	Incorrect, Inadequate or Imprecise Result or Readings	01/09/2023	01/09/2023
							F1905	Device Revision or Replacement	A0908	Incorrect, Inadequate or Imprecise Result or Readings	01/09/2023	01/09/2023
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A0908	Incorrect, Inadequate or Imprecise Result or Readings	01/09/2023	01/09/2023
							F08	Hospitalization or Prolonged Hospitalization	A0908	Incorrect, Inadequate or Imprecise Result or Readings	01/09/2023	01/09/2023
							F1905	Device Revision or Replacement	A0908	Incorrect, Inadequate or Imprecise Result or Readings	01/09/2023	01/09/2023
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A0908	Incorrect, Inadequate or Imprecise Result or Readings	01/09/2023	01/09/2023
							F08	Hospitalization or Prolonged Hospitalization	A0908	Incorrect, Inadequate or Imprecise Result or Readings	01/09/2023	01/09/2023
							F1905	Device Revision or Replacement	A0908	Incorrect, Inadequate or Imprecise Result or Readings	01/09/2023	01/09/2023
14022	2023	All other reportable incidents	35223	Endocardial/interventricular septal pacing lead	CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070909	Oversensing		
							F26	No Health Consequences or Impact	A070909	Oversensing		
14023	2023	All other reportable incidents	47265	Dual-chamber implantable pacemaker, rate-responsive	CT141	Electrodes/Leads and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem	13/03/2023	13/03/2023
									A0908	Incorrect, Inadequate or Imprecise Result or Readings	13/03/2023	13/03/2023
									A1304	Telemetry Discrepancy	13/03/2023	13/03/2023
							F08	Hospitalization or Prolonged Hospitalization	A0705	Battery Problem	13/03/2023	13/03/2023
									A0908	Incorrect, Inadequate or Imprecise Result or Readings	13/03/2023	13/03/2023
									A1304	Telemetry Discrepancy	13/03/2023	13/03/2023
							F1905	Device Revision or Replacement	A0705	Battery Problem	13/03/2023	13/03/2023
									A0908	Incorrect, Inadequate or Imprecise Result or Readings	13/03/2023	13/03/2023
									A1304	Telemetry Discrepancy	13/03/2023	13/03/2023
					CT145	Cardiovascular devices	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem	13/03/2023	13/03/2023
									A0908	Incorrect, Inadequate or Imprecise Result or Readings	13/03/2023	13/03/2023
									A1304	Telemetry Discrepancy	13/03/2023	13/03/2023
							F08	Hospitalization or Prolonged Hospitalization	A0705	Battery Problem	13/03/2023	13/03/2023
									A0908	Incorrect, Inadequate or Imprecise Result or Readings	13/03/2023	13/03/2023

								A1304	Telemetry Discrepancy	13/03/2023	13/03/2023
						F1905	Device Revision or Replacement	A0705	Battery Problem	13/03/2023	13/03/2023
								A0908	Incorrect, Inadequate or Imprecise Result or Readings	13/03/2023	13/03/2023
								A1304	Telemetry Discrepancy	13/03/2023	13/03/2023
					CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem	13/03/2023	13/03/2023
									A0908	Incorrect, Inadequate or Imprecise Result or Readings	13/03/2023	13/03/2023
									A1304	Telemetry Discrepancy	13/03/2023	13/03/2023
							F08	Hospitalization or Prolonged Hospitalization	A0705	Battery Problem	13/03/2023	13/03/2023
									A0908	Incorrect, Inadequate or Imprecise Result or Readings	13/03/2023	13/03/2023
									A1304	Telemetry Discrepancy	13/03/2023	13/03/2023
							F1905	Device Revision or Replacement	A0705	Battery Problem	13/03/2023	13/03/2023
									A0908	Incorrect, Inadequate or Imprecise Result or Readings	13/03/2023	13/03/2023
									A1304	Telemetry Discrepancy	13/03/2023	13/03/2023
					CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem	13/03/2023	13/03/2023
									A0908	Incorrect, Inadequate or Imprecise Result or Readings	13/03/2023	13/03/2023
									A1304	Telemetry Discrepancy	13/03/2023	13/03/2023
							F08	Hospitalization or Prolonged Hospitalization	A0705	Battery Problem	13/03/2023	13/03/2023
									A0908	Incorrect, Inadequate or Imprecise Result or Readings	13/03/2023	13/03/2023
									A1304	Telemetry Discrepancy	13/03/2023	13/03/2023
							F1905	Device Revision or Replacement	A0705	Battery Problem	13/03/2023	13/03/2023
									A0908	Incorrect, Inadequate or Imprecise Result or Readings	13/03/2023	13/03/2023
									A1304	Telemetry Discrepancy	13/03/2023	13/03/2023
					CT233	Surgical	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem	13/03/2023	13/03/2023
									A0908	Incorrect, Inadequate or Imprecise Result or Readings	13/03/2023	13/03/2023
									A1304	Telemetry Discrepancy	13/03/2023	13/03/2023
							F08	Hospitalization or Prolonged Hospitalization	A0705	Battery Problem	13/03/2023	13/03/2023
									A0908	Incorrect, Inadequate or Imprecise Result or Readings	13/03/2023	13/03/2023
									A1304	Telemetry Discrepancy	13/03/2023	13/03/2023
							F1905	Device Revision or Replacement	A0705	Battery Problem	13/03/2023	13/03/2023
									A0908	Incorrect, Inadequate or Imprecise Result or Readings	13/03/2023	13/03/2023
									A1304	Telemetry Discrepancy	13/03/2023	13/03/2023
					CT2406	Implantable	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem	13/03/2023	13/03/2023
									A0908	Incorrect, Inadequate or Imprecise Result or Readings	13/03/2023	13/03/2023
									A1304	Telemetry Discrepancy	13/03/2023	13/03/2023
							F08	Hospitalization or Prolonged Hospitalization	A0705	Battery Problem	13/03/2023	13/03/2023
									A0908	Incorrect, Inadequate or Imprecise Result or Readings	13/03/2023	13/03/2023
									A1304	Telemetry Discrepancy	13/03/2023	13/03/2023
							F1905	Device Revision or Replacement	A0705	Battery Problem	13/03/2023	13/03/2023
									A0908	Incorrect, Inadequate or Imprecise Result or Readings	13/03/2023	13/03/2023
									A1304	Telemetry Discrepancy	13/03/2023	13/03/2023
					CT2407	Active implantable	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem	13/03/2023	13/03/2023
									A0908	Incorrect, Inadequate or Imprecise Result or Readings	13/03/2023	13/03/2023
									A1304	Telemetry Discrepancy	13/03/2023	13/03/2023
							F08	Hospitalization or Prolonged Hospitalization	A0705	Battery Problem	13/03/2023	13/03/2023
									A0908	Incorrect, Inadequate or Imprecise Result or Readings	13/03/2023	13/03/2023
									A1304	Telemetry Discrepancy	13/03/2023	13/03/2023
							F1905	Device Revision or Replacement	A0705	Battery Problem	13/03/2023	13/03/2023
									A0908	Incorrect, Inadequate or Imprecise Result or Readings	13/03/2023	13/03/2023
									A1304	Telemetry Discrepancy	13/03/2023	13/03/2023
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem	13/03/2023	13/03/2023
									A0908	Incorrect, Inadequate or Imprecise Result or Readings	13/03/2023	13/03/2023
									A1304	Telemetry Discrepancy	13/03/2023	13/03/2023
							F08	Hospitalization or Prolonged Hospitalization	A0705	Battery Problem	13/03/2023	13/03/2023
									A0908	Incorrect, Inadequate or Imprecise Result or Readings	13/03/2023	13/03/2023
									A1304	Telemetry Discrepancy	13/03/2023	13/03/2023
							F1905	Device Revision or Replacement	A0705	Battery Problem	13/03/2023	13/03/2023
									A0908	Incorrect, Inadequate or Imprecise Result or Readings	13/03/2023	13/03/2023
									A1304	Telemetry Discrepancy	13/03/2023	13/03/2023
					CT2727	Active implant devices/ancillaries	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem	13/03/2023	13/03/2023
									A0908	Incorrect, Inadequate or Imprecise Result or Readings	13/03/2023	13/03/2023
									A1304	Telemetry Discrepancy	13/03/2023	13/03/2023
							F08	Hospitalization or Prolonged Hospitalization	A0705	Battery Problem	13/03/2023	13/03/2023
									A0908	Incorrect, Inadequate or Imprecise Result or Readings	13/03/2023	13/03/2023
									A1304	Telemetry Discrepancy	13/03/2023	13/03/2023
							F1905	Device Revision or Replacement	A0705	Battery Problem	13/03/2023	13/03/2023
									A0908	Incorrect, Inadequate or Imprecise Result or Readings	13/03/2023	13/03/2023
									A1304	Telemetry Discrepancy	13/03/2023	13/03/2023
					CT2750	Implantable cardiac electrical stimulators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem	13/03/2023	13/03/2023
									A0908	Incorrect, Inadequate or Imprecise Result or Readings	13/03/2023	13/03/2023
									A1304	Telemetry Discrepancy	13/03/2023	13/03/2023

							F08	Hospitalization or Prolonged Hospitalization	A0705	Battery Problem	13/03/2023	13/03/2023
									A0908	Incorrect, Inadequate or Imprecise Result or Readings	13/03/2023	13/03/2023
									A1304	Telemetry Discrepancy	13/03/2023	13/03/2023
							F1905	Device Revision or Replacement	A0705	Battery Problem	13/03/2023	13/03/2023
									A0908	Incorrect, Inadequate or Imprecise Result or Readings	13/03/2023	13/03/2023
									A1304	Telemetry Discrepancy	13/03/2023	13/03/2023
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem	13/03/2023	13/03/2023
									A0908	Incorrect, Inadequate or Imprecise Result or Readings	13/03/2023	13/03/2023
									A1304	Telemetry Discrepancy	13/03/2023	13/03/2023
							F08	Hospitalization or Prolonged Hospitalization	A0705	Battery Problem	13/03/2023	13/03/2023
									A0908	Incorrect, Inadequate or Imprecise Result or Readings	13/03/2023	13/03/2023
									A1304	Telemetry Discrepancy	13/03/2023	13/03/2023
							F1905	Device Revision or Replacement	A0705	Battery Problem	13/03/2023	13/03/2023
									A0908	Incorrect, Inadequate or Imprecise Result or Readings	13/03/2023	13/03/2023
									A1304	Telemetry Discrepancy	13/03/2023	13/03/2023
					CT2985	Active	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem	13/03/2023	13/03/2023
									A0908	Incorrect, Inadequate or Imprecise Result or Readings	13/03/2023	13/03/2023
									A1304	Telemetry Discrepancy	13/03/2023	13/03/2023
							F08	Hospitalization or Prolonged Hospitalization	A0705	Battery Problem	13/03/2023	13/03/2023
									A0908	Incorrect, Inadequate or Imprecise Result or Readings	13/03/2023	13/03/2023
									A1304	Telemetry Discrepancy	13/03/2023	13/03/2023
							F1905	Device Revision or Replacement	A0705	Battery Problem	13/03/2023	13/03/2023
									A0908	Incorrect, Inadequate or Imprecise Result or Readings	13/03/2023	13/03/2023
									A1304	Telemetry Discrepancy	13/03/2023	13/03/2023
					CT3001	Implantable metal	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem	13/03/2023	13/03/2023
									A0908	Incorrect, Inadequate or Imprecise Result or Readings	13/03/2023	13/03/2023
									A1304	Telemetry Discrepancy	13/03/2023	13/03/2023
							F08	Hospitalization or Prolonged Hospitalization	A0705	Battery Problem	13/03/2023	13/03/2023
									A0908	Incorrect, Inadequate or Imprecise Result or Readings	13/03/2023	13/03/2023
									A1304	Telemetry Discrepancy	13/03/2023	13/03/2023
							F1905	Device Revision or Replacement	A0705	Battery Problem	13/03/2023	13/03/2023
									A0908	Incorrect, Inadequate or Imprecise Result or Readings	13/03/2023	13/03/2023
									A1304	Telemetry Discrepancy	13/03/2023	13/03/2023
					CT3002	Implantable synthetic polymer	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem	13/03/2023	13/03/2023
									A0908	Incorrect, Inadequate or Imprecise Result or Readings	13/03/2023	13/03/2023
									A1304	Telemetry Discrepancy	13/03/2023	13/03/2023
							F08	Hospitalization or Prolonged Hospitalization	A0705	Battery Problem	13/03/2023	13/03/2023
									A0908	Incorrect, Inadequate or Imprecise Result or Readings	13/03/2023	13/03/2023
									A1304	Telemetry Discrepancy	13/03/2023	13/03/2023
							F1905	Device Revision or Replacement	A0705	Battery Problem	13/03/2023	13/03/2023
									A0908	Incorrect, Inadequate or Imprecise Result or Readings	13/03/2023	13/03/2023
									A1304	Telemetry Discrepancy	13/03/2023	13/03/2023
					CT3006	Implantable inorganic material	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem	13/03/2023	13/03/2023
									A0908	Incorrect, Inadequate or Imprecise Result or Readings	13/03/2023	13/03/2023
									A1304	Telemetry Discrepancy	13/03/2023	13/03/2023
							F08	Hospitalization or Prolonged Hospitalization	A0705	Battery Problem	13/03/2023	13/03/2023
									A0908	Incorrect, Inadequate or Imprecise Result or Readings	13/03/2023	13/03/2023
									A1304	Telemetry Discrepancy	13/03/2023	13/03/2023
							F1905	Device Revision or Replacement	A0705	Battery Problem	13/03/2023	13/03/2023
									A0908	Incorrect, Inadequate or Imprecise Result or Readings	13/03/2023	13/03/2023
									A1304	Telemetry Discrepancy	13/03/2023	13/03/2023
					CT314	Electrically powered	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem	13/03/2023	13/03/2023
									A0908	Incorrect, Inadequate or Imprecise Result or Readings	13/03/2023	13/03/2023
									A1304	Telemetry Discrepancy	13/03/2023	13/03/2023
							F08	Hospitalization or Prolonged Hospitalization	A0705	Battery Problem	13/03/2023	13/03/2023
									A0908	Incorrect, Inadequate or Imprecise Result or Readings	13/03/2023	13/03/2023
									A1304	Telemetry Discrepancy	13/03/2023	13/03/2023
							F1905	Device Revision or Replacement	A0705	Battery Problem	13/03/2023	13/03/2023
									A0908	Incorrect, Inadequate or Imprecise Result or Readings	13/03/2023	13/03/2023
									A1304	Telemetry Discrepancy	13/03/2023	13/03/2023
					CT321	Long-term surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem	13/03/2023	13/03/2023
									A0908	Incorrect, Inadequate or Imprecise Result or Readings	13/03/2023	13/03/2023
									A1304	Telemetry Discrepancy	13/03/2023	13/03/2023
							F08	Hospitalization or Prolonged Hospitalization	A0705	Battery Problem	13/03/2023	13/03/2023
									A0908	Incorrect, Inadequate or Imprecise Result or Readings	13/03/2023	13/03/2023
									A1304	Telemetry Discrepancy	13/03/2023	13/03/2023
							F1905	Device Revision or Replacement	A0705	Battery Problem	13/03/2023	13/03/2023
									A0908	Incorrect, Inadequate or Imprecise Result or Readings	13/03/2023	13/03/2023
									A1304	Telemetry Discrepancy	13/03/2023	13/03/2023
					CT449	Generators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem	13/03/2023	13/03/2023

									A0908	Incorrect, Inadequate or Imprecise Result or Readings	13/03/2023	13/03/2023
									A1304	Telemetry Discrepancy	13/03/2023	13/03/2023
							F08	Hospitalization or Prolonged Hospitalization	A0705	Battery Problem	13/03/2023	13/03/2023
									A0908	Incorrect, Inadequate or Imprecise Result or Readings	13/03/2023	13/03/2023
									A1304	Telemetry Discrepancy	13/03/2023	13/03/2023
							F1905	Device Revision or Replacement	A0705	Battery Problem	13/03/2023	13/03/2023
									A0908	Incorrect, Inadequate or Imprecise Result or Readings	13/03/2023	13/03/2023
									A1304	Telemetry Discrepancy	13/03/2023	13/03/2023
					CT981	Single use	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem	13/03/2023	13/03/2023
									A0908	Incorrect, Inadequate or Imprecise Result or Readings	13/03/2023	13/03/2023
									A1304	Telemetry Discrepancy	13/03/2023	13/03/2023
							F08	Hospitalization or Prolonged Hospitalization	A0705	Battery Problem	13/03/2023	13/03/2023
									A0908	Incorrect, Inadequate or Imprecise Result or Readings	13/03/2023	13/03/2023
									A1304	Telemetry Discrepancy	13/03/2023	13/03/2023
							F1905	Device Revision or Replacement	A0705	Battery Problem	13/03/2023	13/03/2023
									A0908	Incorrect, Inadequate or Imprecise Result or Readings	13/03/2023	13/03/2023
									A1304	Telemetry Discrepancy	13/03/2023	13/03/2023
					CT983	Surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem	13/03/2023	13/03/2023
									A0908	Incorrect, Inadequate or Imprecise Result or Readings	13/03/2023	13/03/2023
									A1304	Telemetry Discrepancy	13/03/2023	13/03/2023
							F08	Hospitalization or Prolonged Hospitalization	A0705	Battery Problem	13/03/2023	13/03/2023
									A0908	Incorrect, Inadequate or Imprecise Result or Readings	13/03/2023	13/03/2023
									A1304	Telemetry Discrepancy	13/03/2023	13/03/2023
							F1905	Device Revision or Replacement	A0705	Battery Problem	13/03/2023	13/03/2023
									A0908	Incorrect, Inadequate or Imprecise Result or Readings	13/03/2023	13/03/2023
									A1304	Telemetry Discrepancy	13/03/2023	13/03/2023
14024	2023	All other reportable incidents	47265	Dual-chamber implantable pacemaker, rate-responsive			E2403	No Clinical Signs, Symptoms or Conditions	A0908	Incorrect, Inadequate or Imprecise Result or Readings	01/08/2023	01/08/2023
							F08	Hospitalization or Prolonged Hospitalization	A0908	Incorrect, Inadequate or Imprecise Result or Readings	01/08/2023	01/08/2023
							F1905	Device Revision or Replacement	A0908	Incorrect, Inadequate or Imprecise Result or Readings	01/08/2023	01/08/2023
14025	2023	All other reportable incidents	37265	Dual-chamber implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E0601	Arrhythmia	A071201	Failure to Convert Rhythm		
							F23	Unexpected Medical Intervention	A071201	Failure to Convert Rhythm		
14026	2023	All other reportable incidents	47263	Cardiac resynchronization therapy implantable pacemaker	CT1685	Cardiac pacemakers	E011902	Presyncope	A0908	Incorrect, Inadequate or Imprecise Result or Readings	19/06/2023	19/06/2023
							E012207	Twitching	A0908	Incorrect, Inadequate or Imprecise Result or Readings	19/06/2023	19/06/2023
							E1204	Hot Flashes/Flushes	A0908	Incorrect, Inadequate or Imprecise Result or Readings	19/06/2023	19/06/2023
							F08	Hospitalization or Prolonged Hospitalization	A0908	Incorrect, Inadequate or Imprecise Result or Readings	19/06/2023	19/06/2023
							F1905	Device Revision or Replacement	A0908	Incorrect, Inadequate or Imprecise Result or Readings	19/06/2023	19/06/2023
					CT257	Implantable cardiac pacemakers	E011902	Presyncope	A0908	Incorrect, Inadequate or Imprecise Result or Readings	19/06/2023	19/06/2023
							E012207	Twitching	A0908	Incorrect, Inadequate or Imprecise Result or Readings	19/06/2023	19/06/2023
							E1204	Hot Flashes/Flushes	A0908	Incorrect, Inadequate or Imprecise Result or Readings	19/06/2023	19/06/2023
							F08	Hospitalization or Prolonged Hospitalization	A0908	Incorrect, Inadequate or Imprecise Result or Readings	19/06/2023	19/06/2023
							F1905	Device Revision or Replacement	A0908	Incorrect, Inadequate or Imprecise Result or Readings	19/06/2023	19/06/2023
					CT2866	Implantable cardiac electrical stimulators	E011902	Presyncope	A0908	Incorrect, Inadequate or Imprecise Result or Readings	19/06/2023	19/06/2023
							E012207	Twitching	A0908	Incorrect, Inadequate or Imprecise Result or Readings	19/06/2023	19/06/2023
							E1204	Hot Flashes/Flushes	A0908	Incorrect, Inadequate or Imprecise Result or Readings	19/06/2023	19/06/2023
							F08	Hospitalization or Prolonged Hospitalization	A0908	Incorrect, Inadequate or Imprecise Result or Readings	19/06/2023	19/06/2023
							F1905	Device Revision or Replacement	A0908	Incorrect, Inadequate or Imprecise Result or Readings	19/06/2023	19/06/2023
14027	2023	All other reportable incidents	37265	Dual-chamber implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E011903	Syncope/Fainting	A070910	Undersensing		
									A0712	Pacing Problem		
									A071302	Inappropriate/Inadequate Shock/Stimulation		
									A090801	Signal Artifact/Noise		
									A1301	Failure to Read Input Signal		
							E0601	Arrhythmia	A070910	Undersensing		
									A0712	Pacing Problem		
									A071302	Inappropriate/Inadequate Shock/Stimulation		
									A090801	Signal Artifact/Noise		
									A1301	Failure to Read Input Signal		
							E2104	Electric Shock	A070910	Undersensing		
									A0712	Pacing Problem		
									A071302	Inappropriate/Inadequate Shock/Stimulation		
									A090801	Signal Artifact/Noise		
									A1301	Failure to Read Input Signal		
							F08	Hospitalization or Prolonged Hospitalization	A070910	Undersensing		
									A0712	Pacing Problem		
									A071302	Inappropriate/Inadequate Shock/Stimulation		
									A090801	Signal Artifact/Noise		
									A1301	Failure to Read Input Signal		
							F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A070910	Undersensing		

									A0712	Pacing Problem		
									A071302	Inappropriate/Inadequate Shock/Stimulation		
									A090801	Signal Artifact/Noise		
									A1301	Failure to Read Input Signal		
							F2203	Imaging Required	A070910	Undersensing		
									A0712	Pacing Problem		
									A071302	Inappropriate/Inadequate Shock/Stimulation		
									A090801	Signal Artifact/Noise		
									A1301	Failure to Read Input Signal		
14028	2023	All other reportable incidents	37265	Dual-chamber implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem	05/10/2023	05/10/2023
									A070504	Premature Discharge of Battery	05/10/2023	05/10/2023
									A1304	Telemetry Discrepancy	05/10/2023	05/10/2023
							F08	Hospitalization or Prolonged Hospitalization	A0705	Battery Problem	05/10/2023	05/10/2023
									A070504	Premature Discharge of Battery	05/10/2023	05/10/2023
									A1304	Telemetry Discrepancy	05/10/2023	05/10/2023
							F1903	Device Explantation	A0705	Battery Problem	05/10/2023	05/10/2023
									A070504	Premature Discharge of Battery	05/10/2023	05/10/2023
									A1304	Telemetry Discrepancy	05/10/2023	05/10/2023
							F1905	Device Revision or Replacement	A0705	Battery Problem	05/10/2023	05/10/2023
									A070504	Premature Discharge of Battery	05/10/2023	05/10/2023
									A1304	Telemetry Discrepancy	05/10/2023	05/10/2023
14029	2023		47265	Dual-chamber implantable pacemaker, rate-responsive	CT1685	Cardiac pacemakers	F15	Recognized Device or Procedural Complication	A1102	Application Program Problem		
					CT257	Implantable cardiac pacemakers	F15	Recognized Device or Procedural Complication	A1102	Application Program Problem		
					CT2866	Implantable cardiac electrical stimulators	F15	Recognized Device or Procedural Complication	A1102	Application Program Problem		
14030	2023		47265	Dual-chamber implantable pacemaker, rate-responsive	CT141	Electrodes/Leads and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070101	Failure to Capture	11/09/2023	11/09/2023
							F1901	Additional Surgery	A070101	Failure to Capture	11/09/2023	11/09/2023
					CT145	Cardiovascular devices	E2403	No Clinical Signs, Symptoms or Conditions	A070101	Failure to Capture	11/09/2023	11/09/2023
							F1901	Additional Surgery	A070101	Failure to Capture	11/09/2023	11/09/2023
					CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070101	Failure to Capture	11/09/2023	11/09/2023
							F1901	Additional Surgery	A070101	Failure to Capture	11/09/2023	11/09/2023
					CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A070101	Failure to Capture	11/09/2023	11/09/2023
							F1901	Additional Surgery	A070101	Failure to Capture	11/09/2023	11/09/2023
					CT233	Surgical	E2403	No Clinical Signs, Symptoms or Conditions	A070101	Failure to Capture	11/09/2023	11/09/2023
							F1901	Additional Surgery	A070101	Failure to Capture	11/09/2023	11/09/2023
					CT2406	Implantable	E2403	No Clinical Signs, Symptoms or Conditions	A070101	Failure to Capture	11/09/2023	11/09/2023
							F1901	Additional Surgery	A070101	Failure to Capture	11/09/2023	11/09/2023
					CT2407	Active implantable	E2403	No Clinical Signs, Symptoms or Conditions	A070101	Failure to Capture	11/09/2023	11/09/2023
							F1901	Additional Surgery	A070101	Failure to Capture	11/09/2023	11/09/2023
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A070101	Failure to Capture	11/09/2023	11/09/2023
							F1901	Additional Surgery	A070101	Failure to Capture	11/09/2023	11/09/2023
					CT2727	Active implant devices/ancillaries	E2403	No Clinical Signs, Symptoms or Conditions	A070101	Failure to Capture	11/09/2023	11/09/2023
							F1901	Additional Surgery	A070101	Failure to Capture	11/09/2023	11/09/2023
					CT2750	Implantable cardiac electrical stimulators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070101	Failure to Capture	11/09/2023	11/09/2023
							F1901	Additional Surgery	A070101	Failure to Capture	11/09/2023	11/09/2023
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A070101	Failure to Capture	11/09/2023	11/09/2023
							F1901	Additional Surgery	A070101	Failure to Capture	11/09/2023	11/09/2023
					CT2985	Active	E2403	No Clinical Signs, Symptoms or Conditions	A070101	Failure to Capture	11/09/2023	11/09/2023
							F1901	Additional Surgery	A070101	Failure to Capture	11/09/2023	11/09/2023
					CT3001	Implantable metal	E2403	No Clinical Signs, Symptoms or Conditions	A070101	Failure to Capture	11/09/2023	11/09/2023
							F1901	Additional Surgery	A070101	Failure to Capture	11/09/2023	11/09/2023
					CT3002	Implantable synthetic polymer	E2403	No Clinical Signs, Symptoms or Conditions	A070101	Failure to Capture	11/09/2023	11/09/2023
							F1901	Additional Surgery	A070101	Failure to Capture	11/09/2023	11/09/2023
					CT3006	Implantable inorganic material	E2403	No Clinical Signs, Symptoms or Conditions	A070101	Failure to Capture	11/09/2023	11/09/2023
							F1901	Additional Surgery	A070101	Failure to Capture	11/09/2023	11/09/2023
					CT314	Electrically powered	E2403	No Clinical Signs, Symptoms or Conditions	A070101	Failure to Capture	11/09/2023	11/09/2023
							F1901	Additional Surgery	A070101	Failure to Capture	11/09/2023	11/09/2023
					CT321	Long-term surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A070101	Failure to Capture	11/09/2023	11/09/2023
							F1901	Additional Surgery	A070101	Failure to Capture	11/09/2023	11/09/2023
					CT449	Generators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070101	Failure to Capture	11/09/2023	11/09/2023
							F1901	Additional Surgery	A070101	Failure to Capture	11/09/2023	11/09/2023
					CT981	Single use	E2403	No Clinical Signs, Symptoms or Conditions	A070101	Failure to Capture	11/09/2023	11/09/2023
							F1901	Additional Surgery	A070101	Failure to Capture	11/09/2023	11/09/2023
					CT983	Surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A070101	Failure to Capture	11/09/2023	11/09/2023
							F1901	Additional Surgery	A070101	Failure to Capture	11/09/2023	11/09/2023
14031	2023	All other reportable incidents	17882	Physiologic-monitoring defibrillation system	CT1684	Cardiac pacemakers and associated devices	E2401	Insufficient Information	A090811	Unable to Obtain Readings		
							F26	No Health Consequences or Impact	A090811	Unable to Obtain Readings		
14032	2023	All other reportable incidents	63188	Left-ventricular wireless pacing system sensor/transmitter	CT141	Electrodes/Leads and associated devices	F26	No Health Consequences or Impact	A0708	Power Problem		
					CT145	Cardiovascular devices	F26	No Health Consequences or Impact	A0708	Power Problem		
					CT1577	Wireless monitoring system transmitters	F26	No Health Consequences or Impact	A0708	Power Problem		
					CT1684	Cardiac pacemakers and associated devices	F26	No Health Consequences or Impact	A0708	Power Problem		
					CT2317	Transmitters/Receivers/Transceivers	F26	No Health Consequences or Impact	A0708	Power Problem		
					CT233	Surgical	F26	No Health Consequences or Impact	A0708	Power Problem		
					CT2406	Implantable	F26	No Health Consequences or Impact	A0708	Power Problem		
					CT2407	Active implantable	F26	No Health Consequences or Impact	A0708	Power Problem		
					CT255	Transducers and associated devices	F26	No Health Consequences or Impact	A0708	Power Problem		
					CT2726	Physiological monitoring/measuring devices	F26	No Health Consequences or Impact	A0708	Power Problem		
					CT2727	Active implant devices/ancillaries	F26	No Health Consequences or Impact	A0708	Power Problem		
					CT2750	Implantable cardiac electrical stimulators and associated devices	F26	No Health Consequences or Impact	A0708	Power Problem		
					CT2885	Physiological monitoring/measuring system transmitters/receivers/transceivers	F26	No Health Consequences or Impact	A0708	Power Problem		
					CT2985	Active	F26	No Health Consequences or Impact	A0708	Power Problem		
					CT3001	Implantable metal	F26	No Health Consequences or Impact	A0708	Power Problem		
					CT3002	Implantable synthetic polymer	F26	No Health Consequences or Impact	A0708	Power Problem		
					CT3006	Implantable inorganic material	F26	No Health Consequences or Impact	A0708	Power Problem		
					CT314	Electrically powered	F26	No Health Consequences or Impact	A0708	Power Problem		
					CT321	Long-term surgical invasive	F26	No Health Consequences or Impact	A0708	Power Problem		
					CT384	Sensors and associated devices	F26	No Health Consequences or Impact	A0708	Power Problem		

					CT449	Generators and associated devices	F26	No Health Consequences or Impact	A0708	Power Problem		
					CT981	Single use	F26	No Health Consequences or Impact	A0708	Power Problem		
					CT983	Surgical invasive	F26	No Health Consequences or Impact	A0708	Power Problem		
					CT994	Sensors	F26	No Health Consequences or Impact	A0708	Power Problem		
14033	2023	All other reportable incidents	35223	Endocardial/interventricular septal pacing lead	CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A090801	Signal Artifact/Noise	18/08/2023	18/08/2023
							F1908	Prolonged Surgery	A090801	Signal Artifact/Noise	18/08/2023	18/08/2023
14034	2023	All other reportable incidents	17882	Physiologic-monitoring defibrillation system	CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A0713	Defibrillation/Stimulation Problem		
							F27	Problem Identified During Non-Clinical Procedure	A0713	Defibrillation/Stimulation Problem		
14035	2023	All other reportable incidents	17882	Physiologic-monitoring defibrillation system	CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A0719	Unexpected Shutdown		
							F26	No Health Consequences or Impact	A0719	Unexpected Shutdown		
14036	2023	All other reportable incidents	17882	Physiologic-monitoring defibrillation system	CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A07	Electrical /Electronic Property Problem		
							F27	Problem Identified During Non-Clinical Procedure	A07	Electrical /Electronic Property Problem		
14037	2023	All other reportable incidents	17882	Physiologic-monitoring defibrillation system	CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A07	Electrical /Electronic Property Problem		
							F27	Problem Identified During Non-Clinical Procedure	A07	Electrical /Electronic Property Problem		
14038	2023	All other reportable incidents	47265	Dual-chamber implantable pacemaker, rate-responsive	CT141	Electrodes/Leads and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A09	Output Problem		
									A090811	Unable to Obtain Readings		
							F26	No Health Consequences or Impact	A09	Output Problem		
									A090811	Unable to Obtain Readings		
					CT145	Cardiovascular devices	E2403	No Clinical Signs, Symptoms or Conditions	A09	Output Problem		
									A090811	Unable to Obtain Readings		
							F26	No Health Consequences or Impact	A09	Output Problem		
									A090811	Unable to Obtain Readings		
					CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A09	Output Problem		
									A090811	Unable to Obtain Readings		
							F26	No Health Consequences or Impact	A09	Output Problem		
									A090811	Unable to Obtain Readings		
					CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A09	Output Problem		
									A090811	Unable to Obtain Readings		
							F26	No Health Consequences or Impact	A09	Output Problem		
									A090811	Unable to Obtain Readings		
					CT233	Surgical	E2403	No Clinical Signs, Symptoms or Conditions	A09	Output Problem		
									A090811	Unable to Obtain Readings		
							F26	No Health Consequences or Impact	A09	Output Problem		
									A090811	Unable to Obtain Readings		
					CT2406	Implantable	E2403	No Clinical Signs, Symptoms or Conditions	A09	Output Problem		
									A090811	Unable to Obtain Readings		
							F26	No Health Consequences or Impact	A09	Output Problem		
									A090811	Unable to Obtain Readings		
					CT2407	Active implantable	E2403	No Clinical Signs, Symptoms or Conditions	A09	Output Problem		
									A090811	Unable to Obtain Readings		
							F26	No Health Consequences or Impact	A09	Output Problem		
									A090811	Unable to Obtain Readings		
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A09	Output Problem		
									A090811	Unable to Obtain Readings		
							F26	No Health Consequences or Impact	A09	Output Problem		
									A090811	Unable to Obtain Readings		
					CT2727	Active implant devices/ancillaries	E2403	No Clinical Signs, Symptoms or Conditions	A09	Output Problem		
									A090811	Unable to Obtain Readings		
							F26	No Health Consequences or Impact	A09	Output Problem		
									A090811	Unable to Obtain Readings		
					CT2750	Implantable cardiac electrical stimulators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A09	Output Problem		
									A090811	Unable to Obtain Readings		
							F26	No Health Consequences or Impact	A09	Output Problem		
									A090811	Unable to Obtain Readings		
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A09	Output Problem		
									A090811	Unable to Obtain Readings		
							F26	No Health Consequences or Impact	A09	Output Problem		
									A090811	Unable to Obtain Readings		
					CT2985	Active	E2403	No Clinical Signs, Symptoms or Conditions	A09	Output Problem		
									A090811	Unable to Obtain Readings		
							F26	No Health Consequences or Impact	A09	Output Problem		
									A090811	Unable to Obtain Readings		
					CT3001	Implantable metal	E2403	No Clinical Signs, Symptoms or Conditions	A09	Output Problem		
									A090811	Unable to Obtain Readings		
							F26	No Health Consequences or Impact	A09	Output Problem		
									A090811	Unable to Obtain Readings		
					CT3002	Implantable synthetic polymer	E2403	No Clinical Signs, Symptoms or Conditions	A09	Output Problem		
									A090811	Unable to Obtain Readings		
							F26	No Health Consequences or Impact	A09	Output Problem		
									A090811	Unable to Obtain Readings		
					CT3006	Implantable inorganic material	E2403	No Clinical Signs, Symptoms or Conditions	A09	Output Problem		
									A090811	Unable to Obtain Readings		
							F26	No Health Consequences or Impact	A09	Output Problem		
									A090811	Unable to Obtain Readings		
					CT314	Electrically powered	E2403	No Clinical Signs, Symptoms or Conditions	A09	Output Problem		
									A090811	Unable to Obtain Readings		
							F26	No Health Consequences or Impact	A09	Output Problem		
									A090811	Unable to Obtain Readings		
					CT321	Long-term surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A09	Output Problem		
									A090811	Unable to Obtain Readings		
							F26	No Health Consequences or Impact	A09	Output Problem		
									A090811	Unable to Obtain Readings		
					CT449	Generators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A09	Output Problem		
									A090811	Unable to Obtain Readings		
							F26	No Health Consequences or Impact	A09	Output Problem		
									A090811	Unable to Obtain Readings		
					CT981	Single use	E2403	No Clinical Signs, Symptoms or Conditions	A09	Output Problem		
									A090811	Unable to Obtain Readings		

							F26	No Health Consequences or Impact	A09	Output Problem		
									A090811	Unable to Obtain Readings		
					CT983	Surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A09	Output Problem		
									A090811	Unable to Obtain Readings		
							F26	No Health Consequences or Impact	A09	Output Problem		
									A090811	Unable to Obtain Readings		
14039	2023	All other reportable incidents	35223	Endocardial/interventricular septal pacing lead	CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A0722	Impedance Problem		
							F26	No Health Consequences or Impact	A0722	Impedance Problem		
14040	2023	All other reportable incidents	47265	Dual-chamber implantable pacemaker, rate-responsive	CT141	Electrodes/Leads and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	25/07/2023	25/07/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	25/07/2023	25/07/2023
					CT145	Cardiovascular devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	25/07/2023	25/07/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	25/07/2023	25/07/2023
					CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	25/07/2023	25/07/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	25/07/2023	25/07/2023
					CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	25/07/2023	25/07/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	25/07/2023	25/07/2023
					CT233	Surgical	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	25/07/2023	25/07/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	25/07/2023	25/07/2023
					CT2406	Implantable	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	25/07/2023	25/07/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	25/07/2023	25/07/2023
					CT2407	Active implantable	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	25/07/2023	25/07/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	25/07/2023	25/07/2023
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	25/07/2023	25/07/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	25/07/2023	25/07/2023
					CT2727	Active implant devices/ancillaries	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	25/07/2023	25/07/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	25/07/2023	25/07/2023
					CT2750	Implantable cardiac electrical stimulators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	25/07/2023	25/07/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	25/07/2023	25/07/2023
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	25/07/2023	25/07/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	25/07/2023	25/07/2023
					CT2985	Active	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	25/07/2023	25/07/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	25/07/2023	25/07/2023
					CT3001	Implantable metal	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	25/07/2023	25/07/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	25/07/2023	25/07/2023
					CT3002	Implantable synthetic polymer	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	25/07/2023	25/07/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	25/07/2023	25/07/2023
					CT3006	Implantable inorganic material	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	25/07/2023	25/07/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	25/07/2023	25/07/2023
					CT314	Electrically powered	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	25/07/2023	25/07/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	25/07/2023	25/07/2023
					CT321	Long-term surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	25/07/2023	25/07/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	25/07/2023	25/07/2023
					CT449	Generators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	25/07/2023	25/07/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	25/07/2023	25/07/2023
					CT981	Single use	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	25/07/2023	25/07/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	25/07/2023	25/07/2023
					CT983	Surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	25/07/2023	25/07/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	25/07/2023	25/07/2023
14041	2023	All other reportable incidents	47270	Cardiac resynchronization therapy implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A070102	High Capture Threshold		
							F08	Hospitalization or Prolonged Hospitalization	A070102	High Capture Threshold		
							F19	Surgical Intervention	A070102	High Capture Threshold		
14042	2023		47265	Dual-chamber implantable pacemaker, rate-responsive	CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	27/06/2023	27/06/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	27/06/2023	27/06/2023
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	27/06/2023	27/06/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	27/06/2023	27/06/2023
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	27/06/2023	27/06/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	27/06/2023	27/06/2023
14043	2023	All other reportable incidents	47263	Cardiac resynchronization therapy implantable pacemaker	CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A0908	Incorrect, Inadequate or Imprecise Result or Readings	15/09/2023	15/09/2023
							F08	Hospitalization or Prolonged Hospitalization	A0908	Incorrect, Inadequate or Imprecise Result or Readings	15/09/2023	15/09/2023
							F1905	Device Revision or Replacement	A0908	Incorrect, Inadequate or Imprecise Result or Readings	15/09/2023	15/09/2023
					CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A0908	Incorrect, Inadequate or Imprecise Result or Readings	15/09/2023	15/09/2023
							F08	Hospitalization or Prolonged Hospitalization	A0908	Incorrect, Inadequate or Imprecise Result or Readings	15/09/2023	15/09/2023
							F1905	Device Revision or Replacement	A0908	Incorrect, Inadequate or Imprecise Result or Readings	15/09/2023	15/09/2023
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A0908	Incorrect, Inadequate or Imprecise Result or Readings	15/09/2023	15/09/2023
							F08	Hospitalization or Prolonged Hospitalization	A0908	Incorrect, Inadequate or Imprecise Result or Readings	15/09/2023	15/09/2023
							F1905	Device Revision or Replacement	A0908	Incorrect, Inadequate or Imprecise Result or Readings	15/09/2023	15/09/2023
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A0908	Incorrect, Inadequate or Imprecise Result or Readings	15/09/2023	15/09/2023
							F08	Hospitalization or Prolonged Hospitalization	A0908	Incorrect, Inadequate or Imprecise Result or Readings	15/09/2023	15/09/2023
							F1905	Device Revision or Replacement	A0908	Incorrect, Inadequate or Imprecise Result or Readings	15/09/2023	15/09/2023
14044	2023	All other reportable incidents	37265	Dual-chamber implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A071102	Failure to Interrogate	13/09/2023	13/09/2023
							F1905	Device Revision or Replacement	A071102	Failure to Interrogate	13/09/2023	13/09/2023
14045	2023	All other reportable incidents	47265	Dual-chamber implantable pacemaker, rate-responsive	CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
14046	2023	All other reportable incidents	17882	Physiologic-monitoring defibrillation system	CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070803	Failure to Power Up		
							F27	Problem Identified During Non-Clinical Procedure	A070803	Failure to Power Up		

14047	2023	All other reportable incidents	47265	Dual-chamber implantable pacemaker, rate-responsive	CT141	Electrodes/Leads and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	27/09/2023	27/09/2023
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	27/09/2023	27/09/2023
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	27/09/2023	27/09/2023
					CT145	Cardiovascular devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	27/09/2023	27/09/2023
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	27/09/2023	27/09/2023
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	27/09/2023	27/09/2023
					CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	27/09/2023	27/09/2023
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	27/09/2023	27/09/2023
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	27/09/2023	27/09/2023
					CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	27/09/2023	27/09/2023
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	27/09/2023	27/09/2023
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	27/09/2023	27/09/2023
					CT233	Surgical	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	27/09/2023	27/09/2023
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	27/09/2023	27/09/2023
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	27/09/2023	27/09/2023
					CT2406	Implantable	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	27/09/2023	27/09/2023
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	27/09/2023	27/09/2023
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	27/09/2023	27/09/2023
					CT2407	Active implantable	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	27/09/2023	27/09/2023
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	27/09/2023	27/09/2023
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	27/09/2023	27/09/2023
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	27/09/2023	27/09/2023
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	27/09/2023	27/09/2023
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	27/09/2023	27/09/2023
					CT2727	Active implant devices/ancillaries	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	27/09/2023	27/09/2023
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	27/09/2023	27/09/2023
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	27/09/2023	27/09/2023
					CT2750	Implantable cardiac electrical stimulators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	27/09/2023	27/09/2023
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	27/09/2023	27/09/2023
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	27/09/2023	27/09/2023
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	27/09/2023	27/09/2023
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	27/09/2023	27/09/2023
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	27/09/2023	27/09/2023
					CT2985	Active	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	27/09/2023	27/09/2023
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	27/09/2023	27/09/2023
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	27/09/2023	27/09/2023
					CT3001	Implantable metal	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	27/09/2023	27/09/2023
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	27/09/2023	27/09/2023
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	27/09/2023	27/09/2023
					CT3002	Implantable synthetic polymer	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	27/09/2023	27/09/2023
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	27/09/2023	27/09/2023
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	27/09/2023	27/09/2023
					CT3006	Implantable inorganic material	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	27/09/2023	27/09/2023
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	27/09/2023	27/09/2023
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	27/09/2023	27/09/2023
					CT314	Electrically powered	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	27/09/2023	27/09/2023
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	27/09/2023	27/09/2023
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	27/09/2023	27/09/2023
					CT321	Long-term surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	27/09/2023	27/09/2023
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	27/09/2023	27/09/2023
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	27/09/2023	27/09/2023
					CT449	Generators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	27/09/2023	27/09/2023
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	27/09/2023	27/09/2023
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	27/09/2023	27/09/2023
					CT981	Single use	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	27/09/2023	27/09/2023
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	27/09/2023	27/09/2023
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	27/09/2023	27/09/2023
					CT983	Surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	27/09/2023	27/09/2023
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	27/09/2023	27/09/2023
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	27/09/2023	27/09/2023
14048	2023		35223	Endocardial/interventricular septal pacing lead	CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070101	Failure to Capture	11/09/2023	11/09/2023
							F1901	Additional Surgery	A070101	Failure to Capture	11/09/2023	11/09/2023
14049	2023	Death	17882	Physiologic-monitoring defibrillation system	CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A071301	Failure to Deliver Shock/Stimulation		
									A0719	Unexpected Shutdown		
							F02	Death	A071301	Failure to Deliver Shock/Stimulation		
									A0719	Unexpected Shutdown		
14050	2023		47265	Dual-chamber implantable pacemaker, rate-responsive	CT141	Electrodes/Leads and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	07/07/2023	07/07/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	07/07/2023	07/07/2023
					CT145	Cardiovascular devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	07/07/2023	07/07/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	07/07/2023	07/07/2023
					CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	07/07/2023	07/07/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	07/07/2023	07/07/2023
					CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	07/07/2023	07/07/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	07/07/2023	07/07/2023
					CT233	Surgical	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	07/07/2023	07/07/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	07/07/2023	07/07/2023
					CT2406	Implantable	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	07/07/2023	07/07/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	07/07/2023	07/07/2023
					CT2407	Active implantable	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	07/07/2023	07/07/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	07/07/2023	07/07/2023
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	07/07/2023	07/07/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	07/07/2023	07/07/2023
					CT2727	Active implant devices/ancillaries	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	07/07/2023	07/07/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	07/07/2023	07/07/2023
					CT2750	Implantable cardiac electrical stimulators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	07/07/2023	07/07/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	07/07/2023	07/07/2023
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	07/07/2023	07/07/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	07/07/2023	07/07/2023

					CT2985	Active	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	07/07/2023	07/07/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	07/07/2023	07/07/2023
					CT3001	Implantable metal	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	07/07/2023	07/07/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	07/07/2023	07/07/2023
					CT3002	Implantable synthetic polymer	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	07/07/2023	07/07/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	07/07/2023	07/07/2023
					CT3006	Implantable inorganic material	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	07/07/2023	07/07/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	07/07/2023	07/07/2023
					CT314	Electrically powered	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	07/07/2023	07/07/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	07/07/2023	07/07/2023
					CT321	Long-term surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	07/07/2023	07/07/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	07/07/2023	07/07/2023
					CT449	Generators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	07/07/2023	07/07/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	07/07/2023	07/07/2023
					CT981	Single use	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	07/07/2023	07/07/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	07/07/2023	07/07/2023
					CT983	Surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	07/07/2023	07/07/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	07/07/2023	07/07/2023
14051	2023	All other reportable incidents	17882	Physiologic-monitoring defibrillation system	CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A0719	Unexpected Shutdown		
							F26	No Health Consequences or Impact	A0719	Unexpected Shutdown		
14052	2023		18145	Cardiac myoplasty electrical stimulation system	CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A05	Mechanical Problem		
							F23	Unexpected Medical Intervention	A05	Mechanical Problem		
							F26	No Health Consequences or Impact	A05	Mechanical Problem		
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A05	Mechanical Problem		
							F23	Unexpected Medical Intervention	A05	Mechanical Problem		
							F26	No Health Consequences or Impact	A05	Mechanical Problem		
14053	2023	All other reportable incidents	47265	Dual-chamber implantable pacemaker, rate-responsive	CT141	Electrodes/Leads and associated devices	E060101	Asystole	A070909	Oversensing		
									A0712	Pacing Problem		
									A090801	Signal Artifact/Noise		
							F1001	Absence of Treatment	A070909	Oversensing		
									A0712	Pacing Problem		
									A090801	Signal Artifact/Noise		
					CT145	Cardiovascular devices	E060101	Asystole	A070909	Oversensing		
									A0712	Pacing Problem		
									A090801	Signal Artifact/Noise		
							F1001	Absence of Treatment	A070909	Oversensing		
									A0712	Pacing Problem		
									A090801	Signal Artifact/Noise		
					CT1684	Cardiac pacemakers and associated devices	E060101	Asystole	A070909	Oversensing		
									A0712	Pacing Problem		
									A090801	Signal Artifact/Noise		
							F1001	Absence of Treatment	A070909	Oversensing		
									A0712	Pacing Problem		
									A090801	Signal Artifact/Noise		
					CT1685	Cardiac pacemakers	E060101	Asystole	A070909	Oversensing		
									A0712	Pacing Problem		
									A090801	Signal Artifact/Noise		
							F1001	Absence of Treatment	A070909	Oversensing		
									A0712	Pacing Problem		
									A090801	Signal Artifact/Noise		
					CT233	Surgical	E060101	Asystole	A070909	Oversensing		
									A0712	Pacing Problem		
									A090801	Signal Artifact/Noise		
							F1001	Absence of Treatment	A070909	Oversensing		
									A0712	Pacing Problem		
									A090801	Signal Artifact/Noise		
					CT2406	Implantable	E060101	Asystole	A070909	Oversensing		
									A0712	Pacing Problem		
									A090801	Signal Artifact/Noise		
							F1001	Absence of Treatment	A070909	Oversensing		
									A0712	Pacing Problem		
									A090801	Signal Artifact/Noise		
					CT2407	Active implantable	E060101	Asystole	A070909	Oversensing		
									A0712	Pacing Problem		
									A090801	Signal Artifact/Noise		
							F1001	Absence of Treatment	A070909	Oversensing		
									A0712	Pacing Problem		
									A090801	Signal Artifact/Noise		
					CT257	Implantable cardiac pacemakers	E060101	Asystole	A070909	Oversensing		
									A0712	Pacing Problem		
									A090801	Signal Artifact/Noise		
							F1001	Absence of Treatment	A070909	Oversensing		
									A0712	Pacing Problem		
									A090801	Signal Artifact/Noise		
					CT2727	Active implant devices/ancillaries	E060101	Asystole	A070909	Oversensing		
									A0712	Pacing Problem		
									A090801	Signal Artifact/Noise		
							F1001	Absence of Treatment	A070909	Oversensing		
									A0712	Pacing Problem		
									A090801	Signal Artifact/Noise		
					CT2750	Implantable cardiac electrical stimulators and associated devices	E060101	Asystole	A070909	Oversensing		
									A0712	Pacing Problem		
									A090801	Signal Artifact/Noise		
							F1001	Absence of Treatment	A070909	Oversensing		
									A0712	Pacing Problem		
									A090801	Signal Artifact/Noise		
					CT2866	Implantable cardiac electrical stimulators	E060101	Asystole	A070909	Oversensing		
									A0712	Pacing Problem		

									A090801	Signal Artifact/Noise		
							F1001	Absence of Treatment	A070909	Oversensing		
									A0712	Pacing Problem		
									A090801	Signal Artifact/Noise		
					CT2985	Active	E060101	Asystole	A070909	Oversensing		
									A0712	Pacing Problem		
									A090801	Signal Artifact/Noise		
							F1001	Absence of Treatment	A070909	Oversensing		
									A0712	Pacing Problem		
									A090801	Signal Artifact/Noise		
					CT3001	Implantable metal	E060101	Asystole	A070909	Oversensing		
									A0712	Pacing Problem		
									A090801	Signal Artifact/Noise		
							F1001	Absence of Treatment	A070909	Oversensing		
									A0712	Pacing Problem		
									A090801	Signal Artifact/Noise		
					CT3002	Implantable synthetic polymer	E060101	Asystole	A070909	Oversensing		
									A0712	Pacing Problem		
									A090801	Signal Artifact/Noise		
							F1001	Absence of Treatment	A070909	Oversensing		
									A0712	Pacing Problem		
									A090801	Signal Artifact/Noise		
					CT3006	Implantable inorganic material	E060101	Asystole	A070909	Oversensing		
									A0712	Pacing Problem		
									A090801	Signal Artifact/Noise		
							F1001	Absence of Treatment	A070909	Oversensing		
									A0712	Pacing Problem		
									A090801	Signal Artifact/Noise		
					CT314	Electrically powered	E060101	Asystole	A070909	Oversensing		
									A0712	Pacing Problem		
									A090801	Signal Artifact/Noise		
							F1001	Absence of Treatment	A070909	Oversensing		
									A0712	Pacing Problem		
									A090801	Signal Artifact/Noise		
					CT321	Long-term surgical invasive	E060101	Asystole	A070909	Oversensing		
									A0712	Pacing Problem		
									A090801	Signal Artifact/Noise		
							F1001	Absence of Treatment	A070909	Oversensing		
									A0712	Pacing Problem		
									A090801	Signal Artifact/Noise		
					CT449	Generators and associated devices	E060101	Asystole	A070909	Oversensing		
									A0712	Pacing Problem		
									A090801	Signal Artifact/Noise		
							F1001	Absence of Treatment	A070909	Oversensing		
									A0712	Pacing Problem		
									A090801	Signal Artifact/Noise		
					CT981	Single use	E060101	Asystole	A070909	Oversensing		
									A0712	Pacing Problem		
									A090801	Signal Artifact/Noise		
							F1001	Absence of Treatment	A070909	Oversensing		
									A0712	Pacing Problem		
									A090801	Signal Artifact/Noise		
					CT983	Surgical invasive	E060101	Asystole	A070909	Oversensing		
									A0712	Pacing Problem		
									A090801	Signal Artifact/Noise		
							F1001	Absence of Treatment	A070909	Oversensing		
									A0712	Pacing Problem		
									A090801	Signal Artifact/Noise		
14054	2023	All other reportable incidents	36241	Pacemaker/defibrillator lead	CT1684	Cardiac pacemakers and associated devices	E2104	Electric Shock	A070909	Oversensing		
									A071302	Inappropriate/Inadequate Shock/Stimulation		
									A090801	Signal Artifact/Noise		
									A1301	Failure to Read Input Signal		
							F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A070909	Oversensing		
									A071302	Inappropriate/Inadequate Shock/Stimulation		
									A090801	Signal Artifact/Noise		
									A1301	Failure to Read Input Signal		
14055	2023	All other reportable incidents	35852	Single-chamber implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E2104	Electric Shock	A010402	Migration	14/09/2023	14/09/2023
									A070602	Delayed Charge Time	14/09/2023	14/09/2023
									A071302	Inappropriate/Inadequate Shock/Stimulation	14/09/2023	14/09/2023
									A072202	Low Impedance	14/09/2023	14/09/2023
									A0908	Incorrect, Inadequate or Imprecise Result or Readings	14/09/2023	14/09/2023
							E2117	Twiddlers Syndrome	A010402	Migration	14/09/2023	14/09/2023
									A070602	Delayed Charge Time	14/09/2023	14/09/2023
									A071302	Inappropriate/Inadequate Shock/Stimulation	14/09/2023	14/09/2023
									A072202	Low Impedance	14/09/2023	14/09/2023
									A0908	Incorrect, Inadequate or Imprecise Result or Readings	14/09/2023	14/09/2023
							F08	Hospitalization or Prolonged Hospitalization	A010402	Migration	14/09/2023	14/09/2023
									A070602	Delayed Charge Time	14/09/2023	14/09/2023
									A071302	Inappropriate/Inadequate Shock/Stimulation	14/09/2023	14/09/2023
									A072202	Low Impedance	14/09/2023	14/09/2023
									A0908	Incorrect, Inadequate or Imprecise Result or Readings	14/09/2023	14/09/2023
							F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A010402	Migration	14/09/2023	14/09/2023
									A070602	Delayed Charge Time	14/09/2023	14/09/2023
									A071302	Inappropriate/Inadequate Shock/Stimulation	14/09/2023	14/09/2023
									A072202	Low Impedance	14/09/2023	14/09/2023

								A0908	Incorrect, Inadequate or Imprecise Result or Readings	14/09/2023	14/09/2023
						F1903	Device Explantation	A010402	Migration	14/09/2023	14/09/2023
								A070602	Delayed Charge Time	14/09/2023	14/09/2023
								A071302	Inappropriate/Inadequate Shock/Stimulation	14/09/2023	14/09/2023
								A072202	Low Impedance	14/09/2023	14/09/2023
								A0908	Incorrect, Inadequate or Imprecise Result or Readings	14/09/2023	14/09/2023
						F2203	Imaging Required	A010402	Migration	14/09/2023	14/09/2023
								A070602	Delayed Charge Time	14/09/2023	14/09/2023
								A071302	Inappropriate/Inadequate Shock/Stimulation	14/09/2023	14/09/2023
								A072202	Low Impedance	14/09/2023	14/09/2023
								A0908	Incorrect, Inadequate or Imprecise Result or Readings	14/09/2023	14/09/2023
						F23	Unexpected Medical Intervention	A010402	Migration	14/09/2023	14/09/2023
								A070602	Delayed Charge Time	14/09/2023	14/09/2023
								A071302	Inappropriate/Inadequate Shock/Stimulation	14/09/2023	14/09/2023
								A072202	Low Impedance	14/09/2023	14/09/2023
								A0908	Incorrect, Inadequate or Imprecise Result or Readings	14/09/2023	14/09/2023
14056	2023	All other reportable incidents	35223	Endocardial/interventricular septal pacing lead	CT1684	Cardiac pacemakers and associated devices	E233001	Chest Pain	A051201	Device Dislodged or Dislocated		
									A0701	Capturing Problem		
									A070103	Intermittent Capture		
									A070910	Undersensing		
							F12	Serious Injury/ Illness/ Impairment	A051201	Device Dislodged or Dislocated		
									A0701	Capturing Problem		
									A070103	Intermittent Capture		
									A070910	Undersensing		
							F23	Unexpected Medical Intervention	A051201	Device Dislodged or Dislocated		
									A0701	Capturing Problem		
									A070103	Intermittent Capture		
									A070910	Undersensing		
14057	2023	All other reportable incidents	47265	Dual-chamber implantable pacemaker, rate-responsive	CT141	Electrodes/Leads and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A0908	Incorrect, Inadequate or Imprecise Result or Readings	20/03/2023	20/03/2023
							F08	Hospitalization or Prolonged Hospitalization	A0908	Incorrect, Inadequate or Imprecise Result or Readings	20/03/2023	20/03/2023
							F1905	Device Revision or Replacement	A0908	Incorrect, Inadequate or Imprecise Result or Readings	20/03/2023	20/03/2023
					CT145	Cardiovascular devices	E2403	No Clinical Signs, Symptoms or Conditions	A0908	Incorrect, Inadequate or Imprecise Result or Readings	20/03/2023	20/03/2023
							F08	Hospitalization or Prolonged Hospitalization	A0908	Incorrect, Inadequate or Imprecise Result or Readings	20/03/2023	20/03/2023
							F1905	Device Revision or Replacement	A0908	Incorrect, Inadequate or Imprecise Result or Readings	20/03/2023	20/03/2023
					CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A0908	Incorrect, Inadequate or Imprecise Result or Readings	20/03/2023	20/03/2023
							F08	Hospitalization or Prolonged Hospitalization	A0908	Incorrect, Inadequate or Imprecise Result or Readings	20/03/2023	20/03/2023
							F1905	Device Revision or Replacement	A0908	Incorrect, Inadequate or Imprecise Result or Readings	20/03/2023	20/03/2023
					CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A0908	Incorrect, Inadequate or Imprecise Result or Readings	20/03/2023	20/03/2023
							F08	Hospitalization or Prolonged Hospitalization	A0908	Incorrect, Inadequate or Imprecise Result or Readings	20/03/2023	20/03/2023
							F1905	Device Revision or Replacement	A0908	Incorrect, Inadequate or Imprecise Result or Readings	20/03/2023	20/03/2023
					CT233	Surgical	E2403	No Clinical Signs, Symptoms or Conditions	A0908	Incorrect, Inadequate or Imprecise Result or Readings	20/03/2023	20/03/2023
							F08	Hospitalization or Prolonged Hospitalization	A0908	Incorrect, Inadequate or Imprecise Result or Readings	20/03/2023	20/03/2023
							F1905	Device Revision or Replacement	A0908	Incorrect, Inadequate or Imprecise Result or Readings	20/03/2023	20/03/2023
					CT2406	Implantable	E2403	No Clinical Signs, Symptoms or Conditions	A0908	Incorrect, Inadequate or Imprecise Result or Readings	20/03/2023	20/03/2023
							F08	Hospitalization or Prolonged Hospitalization	A0908	Incorrect, Inadequate or Imprecise Result or Readings	20/03/2023	20/03/2023
							F1905	Device Revision or Replacement	A0908	Incorrect, Inadequate or Imprecise Result or Readings	20/03/2023	20/03/2023
					CT2407	Active implantable	E2403	No Clinical Signs, Symptoms or Conditions	A0908	Incorrect, Inadequate or Imprecise Result or Readings	20/03/2023	20/03/2023
							F08	Hospitalization or Prolonged Hospitalization	A0908	Incorrect, Inadequate or Imprecise Result or Readings	20/03/2023	20/03/2023
							F1905	Device Revision or Replacement	A0908	Incorrect, Inadequate or Imprecise Result or Readings	20/03/2023	20/03/2023
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A0908	Incorrect, Inadequate or Imprecise Result or Readings	20/03/2023	20/03/2023
							F08	Hospitalization or Prolonged Hospitalization	A0908	Incorrect, Inadequate or Imprecise Result or Readings	20/03/2023	20/03/2023
							F1905	Device Revision or Replacement	A0908	Incorrect, Inadequate or Imprecise Result or Readings	20/03/2023	20/03/2023
					CT2727	Active implant devices/ancillaries	E2403	No Clinical Signs, Symptoms or Conditions	A0908	Incorrect, Inadequate or Imprecise Result or Readings	20/03/2023	20/03/2023
							F08	Hospitalization or Prolonged Hospitalization	A0908	Incorrect, Inadequate or Imprecise Result or Readings	20/03/2023	20/03/2023
							F1905	Device Revision or Replacement	A0908	Incorrect, Inadequate or Imprecise Result or Readings	20/03/2023	20/03/2023
					CT2750	Implantable cardiac electrical stimulators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A0908	Incorrect, Inadequate or Imprecise Result or Readings	20/03/2023	20/03/2023
							F08	Hospitalization or Prolonged Hospitalization	A0908	Incorrect, Inadequate or Imprecise Result or Readings	20/03/2023	20/03/2023
							F1905	Device Revision or Replacement	A0908	Incorrect, Inadequate or Imprecise Result or Readings	20/03/2023	20/03/2023
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A0908	Incorrect, Inadequate or Imprecise Result or Readings	20/03/2023	20/03/2023
							F08	Hospitalization or Prolonged Hospitalization	A0908	Incorrect, Inadequate or Imprecise Result or Readings	20/03/2023	20/03/2023

							F1905	Device Revision or Replacement	A0908	Incorrect, Inadequate or Imprecise Result or Readings	20/03/2023	20/03/2023
					CT2985	Active	E2403	No Clinical Signs, Symptoms or Conditions	A0908	Incorrect, Inadequate or Imprecise Result or Readings	20/03/2023	20/03/2023
							F08	Hospitalization or Prolonged Hospitalization	A0908	Incorrect, Inadequate or Imprecise Result or Readings	20/03/2023	20/03/2023
							F1905	Device Revision or Replacement	A0908	Incorrect, Inadequate or Imprecise Result or Readings	20/03/2023	20/03/2023
					CT3001	Implantable metal	E2403	No Clinical Signs, Symptoms or Conditions	A0908	Incorrect, Inadequate or Imprecise Result or Readings	20/03/2023	20/03/2023
							F08	Hospitalization or Prolonged Hospitalization	A0908	Incorrect, Inadequate or Imprecise Result or Readings	20/03/2023	20/03/2023
							F1905	Device Revision or Replacement	A0908	Incorrect, Inadequate or Imprecise Result or Readings	20/03/2023	20/03/2023
					CT3002	Implantable synthetic polymer	E2403	No Clinical Signs, Symptoms or Conditions	A0908	Incorrect, Inadequate or Imprecise Result or Readings	20/03/2023	20/03/2023
							F08	Hospitalization or Prolonged Hospitalization	A0908	Incorrect, Inadequate or Imprecise Result or Readings	20/03/2023	20/03/2023
							F1905	Device Revision or Replacement	A0908	Incorrect, Inadequate or Imprecise Result or Readings	20/03/2023	20/03/2023
					CT3006	Implantable inorganic material	E2403	No Clinical Signs, Symptoms or Conditions	A0908	Incorrect, Inadequate or Imprecise Result or Readings	20/03/2023	20/03/2023
							F08	Hospitalization or Prolonged Hospitalization	A0908	Incorrect, Inadequate or Imprecise Result or Readings	20/03/2023	20/03/2023
							F1905	Device Revision or Replacement	A0908	Incorrect, Inadequate or Imprecise Result or Readings	20/03/2023	20/03/2023
					CT314	Electrically powered	E2403	No Clinical Signs, Symptoms or Conditions	A0908	Incorrect, Inadequate or Imprecise Result or Readings	20/03/2023	20/03/2023
							F08	Hospitalization or Prolonged Hospitalization	A0908	Incorrect, Inadequate or Imprecise Result or Readings	20/03/2023	20/03/2023
							F1905	Device Revision or Replacement	A0908	Incorrect, Inadequate or Imprecise Result or Readings	20/03/2023	20/03/2023
					CT321	Long-term surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A0908	Incorrect, Inadequate or Imprecise Result or Readings	20/03/2023	20/03/2023
							F08	Hospitalization or Prolonged Hospitalization	A0908	Incorrect, Inadequate or Imprecise Result or Readings	20/03/2023	20/03/2023
							F1905	Device Revision or Replacement	A0908	Incorrect, Inadequate or Imprecise Result or Readings	20/03/2023	20/03/2023
					CT449	Generators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A0908	Incorrect, Inadequate or Imprecise Result or Readings	20/03/2023	20/03/2023
							F08	Hospitalization or Prolonged Hospitalization	A0908	Incorrect, Inadequate or Imprecise Result or Readings	20/03/2023	20/03/2023
							F1905	Device Revision or Replacement	A0908	Incorrect, Inadequate or Imprecise Result or Readings	20/03/2023	20/03/2023
					CT981	Single use	E2403	No Clinical Signs, Symptoms or Conditions	A0908	Incorrect, Inadequate or Imprecise Result or Readings	20/03/2023	20/03/2023
							F08	Hospitalization or Prolonged Hospitalization	A0908	Incorrect, Inadequate or Imprecise Result or Readings	20/03/2023	20/03/2023
							F1905	Device Revision or Replacement	A0908	Incorrect, Inadequate or Imprecise Result or Readings	20/03/2023	20/03/2023
					CT983	Surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A0908	Incorrect, Inadequate or Imprecise Result or Readings	20/03/2023	20/03/2023
							F08	Hospitalization or Prolonged Hospitalization	A0908	Incorrect, Inadequate or Imprecise Result or Readings	20/03/2023	20/03/2023
							F1905	Device Revision or Replacement	A0908	Incorrect, Inadequate or Imprecise Result or Readings	20/03/2023	20/03/2023
14058	2023	All other reportable incidents	35852	Single-chamber implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	31/10/2023	31/10/2023
									A0908	Incorrect, Inadequate or Imprecise Result or Readings	31/10/2023	31/10/2023
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	31/10/2023	31/10/2023
									A0908	Incorrect, Inadequate or Imprecise Result or Readings	31/10/2023	31/10/2023
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	31/10/2023	31/10/2023
									A0908	Incorrect, Inadequate or Imprecise Result or Readings	31/10/2023	31/10/2023
14059	2023	All other reportable incidents	36241	Pacemaker/defibrillator lead	CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A072201	High Impedance		
							F26	No Health Consequences or Impact	A072201	High Impedance		
14060	2023	All other reportable incidents	35852	Single-chamber implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem	13/10/2023	13/10/2023
									A070504	Premature Discharge of Battery	13/10/2023	13/10/2023
									A0908	Incorrect, Inadequate or Imprecise Result or Readings	13/10/2023	13/10/2023
							F08	Hospitalization or Prolonged Hospitalization	A0705	Battery Problem	13/10/2023	13/10/2023
									A070504	Premature Discharge of Battery	13/10/2023	13/10/2023
									A0908	Incorrect, Inadequate or Imprecise Result or Readings	13/10/2023	13/10/2023
							F1905	Device Revision or Replacement	A0705	Battery Problem	13/10/2023	13/10/2023
									A070504	Premature Discharge of Battery	13/10/2023	13/10/2023
									A0908	Incorrect, Inadequate or Imprecise Result or Readings	13/10/2023	13/10/2023
14061	2023		47265	Dual-chamber implantable pacemaker, rate-responsive	CT141	Electrodes/Leads and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	20/09/2023	20/09/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	20/09/2023	20/09/2023
					CT145	Cardiovascular devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	20/09/2023	20/09/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	20/09/2023	20/09/2023
					CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	20/09/2023	20/09/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	20/09/2023	20/09/2023
					CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	20/09/2023	20/09/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	20/09/2023	20/09/2023
					CT233	Surgical	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	20/09/2023	20/09/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	20/09/2023	20/09/2023
					CT2406	Implantable	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	20/09/2023	20/09/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	20/09/2023	20/09/2023
					CT2407	Active implantable	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	20/09/2023	20/09/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	20/09/2023	20/09/2023
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	20/09/2023	20/09/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	20/09/2023	20/09/2023
					CT2727	Active implant devices/ancillaries	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	20/09/2023	20/09/2023

							F1901	Additional Surgery	A24	No Apparent Device Problem	20/09/2023	20/09/2023
					CT2750	Implantable cardiac electrical stimulators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	20/09/2023	20/09/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	20/09/2023	20/09/2023
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	20/09/2023	20/09/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	20/09/2023	20/09/2023
					CT2985	Active	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	20/09/2023	20/09/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	20/09/2023	20/09/2023
					CT3001	Implantable metal	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	20/09/2023	20/09/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	20/09/2023	20/09/2023
					CT3002	Implantable synthetic polymer	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	20/09/2023	20/09/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	20/09/2023	20/09/2023
					CT3006	Implantable inorganic material	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	20/09/2023	20/09/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	20/09/2023	20/09/2023
					CT314	Electrically powered	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	20/09/2023	20/09/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	20/09/2023	20/09/2023
					CT321	Long-term surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	20/09/2023	20/09/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	20/09/2023	20/09/2023
					CT449	Generators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	20/09/2023	20/09/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	20/09/2023	20/09/2023
					CT981	Single use	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	20/09/2023	20/09/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	20/09/2023	20/09/2023
					CT983	Surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	20/09/2023	20/09/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	20/09/2023	20/09/2023
14062	2023	All other reportable incidents	35852	Single-chamber implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	14/06/2023	14/06/2023
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	14/06/2023	14/06/2023
							F12	Serious Injury/ Illness/ Impairment	A070504	Premature Discharge of Battery	14/06/2023	14/06/2023
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	14/06/2023	14/06/2023
14063	2023	All other reportable incidents	47265	Dual-chamber implantable pacemaker, rate-responsive	CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	31/05/2023	31/05/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	31/05/2023	31/05/2023
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	31/05/2023	31/05/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	31/05/2023	31/05/2023
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	31/05/2023	31/05/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	31/05/2023	31/05/2023
14064	2023	All other reportable incidents	17882	Physiologic-monitoring defibrillation system	CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A0908	Incorrect, Inadequate or Imprecise Result or Readings		
							F27	Problem Identified During Non-Clinical Procedure	A0908	Incorrect, Inadequate or Imprecise Result or Readings		
14065	2023	All other reportable incidents	37265	Dual-chamber implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
14066	2023	All other reportable incidents	47270	Cardiac resynchronization therapy implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E0623	Unspecified Heart Problem	A070909	Oversensing		
									A090801	Signal Artifact/Noise		
							F26	No Health Consequences or Impact	A070909	Oversensing		
									A090801	Signal Artifact/Noise		
14067	2023	All other reportable incidents	63188	Left-ventricular wireless pacing system sensor/transmitter	CT141	Electrodes/Leads and associated devices			A0708	Power Problem		
					CT145	Cardiovascular devices			A0708	Power Problem		
					CT1577	Wireless monitoring system transmitters			A0708	Power Problem		
					CT1684	Cardiac pacemakers and associated devices			A0708	Power Problem		
					CT2317	Transmitters/Receivers/Transceivers			A0708	Power Problem		
					CT233	Surgical			A0708	Power Problem		
					CT2406	Implantable			A0708	Power Problem		
					CT2407	Active implantable			A0708	Power Problem		
					CT255	Transducers and associated devices			A0708	Power Problem		
					CT2726	Physiological monitoring/measuring devices			A0708	Power Problem		
					CT2727	Active implant devices/ancillaries			A0708	Power Problem		
					CT2750	Implantable cardiac electrical stimulators and associated devices			A0708	Power Problem		
					CT2885	Physiological monitoring/measuring system transmitters/receivers/transceivers			A0708	Power Problem		
					CT2985	Active			A0708	Power Problem		
					CT3001	Implantable metal			A0708	Power Problem		
					CT3002	Implantable synthetic polymer			A0708	Power Problem		
					CT3006	Implantable inorganic material			A0708	Power Problem		
					CT314	Electrically powered			A0708	Power Problem		
					CT321	Long-term surgical invasive			A0708	Power Problem		
					CT384	Sensors and associated devices			A0708	Power Problem		
					CT449	Generators and associated devices			A0708	Power Problem		
					CT981	Single use			A0708	Power Problem		
					CT983	Surgical invasive			A0708	Power Problem		
					CT994	Sensors			A0708	Power Problem		
14068	2023	All other reportable incidents	47265	Dual-chamber implantable pacemaker, rate-responsive	CT141	Electrodes/Leads and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A0908	Incorrect, Inadequate or Imprecise Result or Readings	27/09/2023	27/09/2023
							F08	Hospitalization or Prolonged Hospitalization	A0908	Incorrect, Inadequate or Imprecise Result or Readings	27/09/2023	27/09/2023
							F1905	Device Revision or Replacement	A0908	Incorrect, Inadequate or Imprecise Result or Readings	27/09/2023	27/09/2023
					CT145	Cardiovascular devices	E2403	No Clinical Signs, Symptoms or Conditions	A0908	Incorrect, Inadequate or Imprecise Result or Readings	27/09/2023	27/09/2023
							F08	Hospitalization or Prolonged Hospitalization	A0908	Incorrect, Inadequate or Imprecise Result or Readings	27/09/2023	27/09/2023
							F1905	Device Revision or Replacement	A0908	Incorrect, Inadequate or Imprecise Result or Readings	27/09/2023	27/09/2023
					CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A0908	Incorrect, Inadequate or Imprecise Result or Readings	27/09/2023	27/09/2023
							F08	Hospitalization or Prolonged Hospitalization	A0908	Incorrect, Inadequate or Imprecise Result or Readings	27/09/2023	27/09/2023
							F1905	Device Revision or Replacement	A0908	Incorrect, Inadequate or Imprecise Result or Readings	27/09/2023	27/09/2023
					CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A0908	Incorrect, Inadequate or Imprecise Result or Readings	27/09/2023	27/09/2023
							F08	Hospitalization or Prolonged Hospitalization	A0908	Incorrect, Inadequate or Imprecise Result or Readings	27/09/2023	27/09/2023

							F1905	Device Revision or Replacement	A0908	Incorrect, Inadequate or Imprecise Result or Readings	27/09/2023	27/09/2023
					CT233	Surgical	E2403	No Clinical Signs, Symptoms or Conditions	A0908	Incorrect, Inadequate or Imprecise Result or Readings	27/09/2023	27/09/2023
							F08	Hospitalization or Prolonged Hospitalization	A0908	Incorrect, Inadequate or Imprecise Result or Readings	27/09/2023	27/09/2023
							F1905	Device Revision or Replacement	A0908	Incorrect, Inadequate or Imprecise Result or Readings	27/09/2023	27/09/2023
					CT2406	Implantable	E2403	No Clinical Signs, Symptoms or Conditions	A0908	Incorrect, Inadequate or Imprecise Result or Readings	27/09/2023	27/09/2023
							F08	Hospitalization or Prolonged Hospitalization	A0908	Incorrect, Inadequate or Imprecise Result or Readings	27/09/2023	27/09/2023
							F1905	Device Revision or Replacement	A0908	Incorrect, Inadequate or Imprecise Result or Readings	27/09/2023	27/09/2023
					CT2407	Active implantable	E2403	No Clinical Signs, Symptoms or Conditions	A0908	Incorrect, Inadequate or Imprecise Result or Readings	27/09/2023	27/09/2023
							F08	Hospitalization or Prolonged Hospitalization	A0908	Incorrect, Inadequate or Imprecise Result or Readings	27/09/2023	27/09/2023
							F1905	Device Revision or Replacement	A0908	Incorrect, Inadequate or Imprecise Result or Readings	27/09/2023	27/09/2023
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A0908	Incorrect, Inadequate or Imprecise Result or Readings	27/09/2023	27/09/2023
							F08	Hospitalization or Prolonged Hospitalization	A0908	Incorrect, Inadequate or Imprecise Result or Readings	27/09/2023	27/09/2023
							F1905	Device Revision or Replacement	A0908	Incorrect, Inadequate or Imprecise Result or Readings	27/09/2023	27/09/2023
					CT2727	Active implant devices/ancillaries	E2403	No Clinical Signs, Symptoms or Conditions	A0908	Incorrect, Inadequate or Imprecise Result or Readings	27/09/2023	27/09/2023
							F08	Hospitalization or Prolonged Hospitalization	A0908	Incorrect, Inadequate or Imprecise Result or Readings	27/09/2023	27/09/2023
							F1905	Device Revision or Replacement	A0908	Incorrect, Inadequate or Imprecise Result or Readings	27/09/2023	27/09/2023
					CT2750	Implantable cardiac electrical stimulators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A0908	Incorrect, Inadequate or Imprecise Result or Readings	27/09/2023	27/09/2023
							F08	Hospitalization or Prolonged Hospitalization	A0908	Incorrect, Inadequate or Imprecise Result or Readings	27/09/2023	27/09/2023
							F1905	Device Revision or Replacement	A0908	Incorrect, Inadequate or Imprecise Result or Readings	27/09/2023	27/09/2023
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A0908	Incorrect, Inadequate or Imprecise Result or Readings	27/09/2023	27/09/2023
							F08	Hospitalization or Prolonged Hospitalization	A0908	Incorrect, Inadequate or Imprecise Result or Readings	27/09/2023	27/09/2023
							F1905	Device Revision or Replacement	A0908	Incorrect, Inadequate or Imprecise Result or Readings	27/09/2023	27/09/2023
					CT2985	Active	E2403	No Clinical Signs, Symptoms or Conditions	A0908	Incorrect, Inadequate or Imprecise Result or Readings	27/09/2023	27/09/2023
							F08	Hospitalization or Prolonged Hospitalization	A0908	Incorrect, Inadequate or Imprecise Result or Readings	27/09/2023	27/09/2023
							F1905	Device Revision or Replacement	A0908	Incorrect, Inadequate or Imprecise Result or Readings	27/09/2023	27/09/2023
					CT3001	Implantable metal	E2403	No Clinical Signs, Symptoms or Conditions	A0908	Incorrect, Inadequate or Imprecise Result or Readings	27/09/2023	27/09/2023
							F08	Hospitalization or Prolonged Hospitalization	A0908	Incorrect, Inadequate or Imprecise Result or Readings	27/09/2023	27/09/2023
							F1905	Device Revision or Replacement	A0908	Incorrect, Inadequate or Imprecise Result or Readings	27/09/2023	27/09/2023
					CT3002	Implantable synthetic polymer	E2403	No Clinical Signs, Symptoms or Conditions	A0908	Incorrect, Inadequate or Imprecise Result or Readings	27/09/2023	27/09/2023
							F08	Hospitalization or Prolonged Hospitalization	A0908	Incorrect, Inadequate or Imprecise Result or Readings	27/09/2023	27/09/2023
							F1905	Device Revision or Replacement	A0908	Incorrect, Inadequate or Imprecise Result or Readings	27/09/2023	27/09/2023
					CT3006	Implantable inorganic material	E2403	No Clinical Signs, Symptoms or Conditions	A0908	Incorrect, Inadequate or Imprecise Result or Readings	27/09/2023	27/09/2023
							F08	Hospitalization or Prolonged Hospitalization	A0908	Incorrect, Inadequate or Imprecise Result or Readings	27/09/2023	27/09/2023
							F1905	Device Revision or Replacement	A0908	Incorrect, Inadequate or Imprecise Result or Readings	27/09/2023	27/09/2023
					CT314	Electrically powered	E2403	No Clinical Signs, Symptoms or Conditions	A0908	Incorrect, Inadequate or Imprecise Result or Readings	27/09/2023	27/09/2023
							F08	Hospitalization or Prolonged Hospitalization	A0908	Incorrect, Inadequate or Imprecise Result or Readings	27/09/2023	27/09/2023
							F1905	Device Revision or Replacement	A0908	Incorrect, Inadequate or Imprecise Result or Readings	27/09/2023	27/09/2023
					CT321	Long-term surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A0908	Incorrect, Inadequate or Imprecise Result or Readings	27/09/2023	27/09/2023
							F08	Hospitalization or Prolonged Hospitalization	A0908	Incorrect, Inadequate or Imprecise Result or Readings	27/09/2023	27/09/2023
							F1905	Device Revision or Replacement	A0908	Incorrect, Inadequate or Imprecise Result or Readings	27/09/2023	27/09/2023
					CT449	Generators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A0908	Incorrect, Inadequate or Imprecise Result or Readings	27/09/2023	27/09/2023
							F08	Hospitalization or Prolonged Hospitalization	A0908	Incorrect, Inadequate or Imprecise Result or Readings	27/09/2023	27/09/2023
							F1905	Device Revision or Replacement	A0908	Incorrect, Inadequate or Imprecise Result or Readings	27/09/2023	27/09/2023
					CT981	Single use	E2403	No Clinical Signs, Symptoms or Conditions	A0908	Incorrect, Inadequate or Imprecise Result or Readings	27/09/2023	27/09/2023
							F08	Hospitalization or Prolonged Hospitalization	A0908	Incorrect, Inadequate or Imprecise Result or Readings	27/09/2023	27/09/2023
							F1905	Device Revision or Replacement	A0908	Incorrect, Inadequate or Imprecise Result or Readings	27/09/2023	27/09/2023
					CT983	Surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A0908	Incorrect, Inadequate or Imprecise Result or Readings	27/09/2023	27/09/2023
							F08	Hospitalization or Prolonged Hospitalization	A0908	Incorrect, Inadequate or Imprecise Result or Readings	27/09/2023	27/09/2023
							F1905	Device Revision or Replacement	A0908	Incorrect, Inadequate or Imprecise Result or Readings	27/09/2023	27/09/2023
14069	2023	All other reportable incidents	35223	Endocardial/interventricular septal pacing lead	CT1684	Cardiac pacemakers and associated devices	E0514	Thrombosis/Thrombus	A05	Mechanical Problem		

									A0709	Device Sensing Problem		
									A0722	Impedance Problem		
									A072201	High Impedance		
							E060102	Atrial Fibrillation	A05	Mechanical Problem		
									A0709	Device Sensing Problem		
									A0722	Impedance Problem		
									A072201	High Impedance		
							E0717	Dyspnea	A05	Mechanical Problem		
									A0709	Device Sensing Problem		
									A0722	Impedance Problem		
									A072201	High Impedance		
							F12	Serious Injury/ Illness/ Impairment	A05	Mechanical Problem		
									A0709	Device Sensing Problem		
									A0722	Impedance Problem		
									A072201	High Impedance		
							F2303	Medication Required	A05	Mechanical Problem		
									A0709	Device Sensing Problem		
									A0722	Impedance Problem		
									A072201	High Impedance		
14070	2023	All other reportable incidents	35223	Endocardial/interventricular septal pacing lead	CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A051201	Device Dislodged or Dislocated		
									A0709	Device Sensing Problem		
									A070909	Oversensing		
									A0722	Impedance Problem		
							F26	No Health Consequences or Impact	A051201	Device Dislodged or Dislocated		
									A0709	Device Sensing Problem		
									A070909	Oversensing		
									A0722	Impedance Problem		
14071	2023				CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	20/09/2023	
							F19	Surgical Intervention	A070504	Premature Discharge of Battery	20/09/2023	
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	20/09/2023	
							F21	Unexpected Deterioration	A070504	Premature Discharge of Battery	20/09/2023	
14072	2023	All other reportable incidents	17882	Physiologic-monitoring defibrillation system	CT1684	Cardiac pacemakers and associated devices	E0601	Arrhythmia	A0707	Failure to Discharge		
							F05	Delay to Treatment/ Therapy	A0707	Failure to Discharge		
14073	2023	All other reportable incidents	47205	Cardiac pulse generator programmer	CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A13	Communication or Transmission Problem		
							F23	Unexpected Medical Intervention	A13	Communication or Transmission Problem		
14074	2023	All other reportable incidents	17882	Physiologic-monitoring defibrillation system	CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A0401	Break		
							F27	Problem Identified During Non-Clinical Procedure	A0401	Break		
14075	2023	All other reportable incidents	47267	Single-chamber implantable pacemaker, rate-responsive	CT141	Electrodes/Leads and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	13/09/2023	13/09/2023
									A071102	Failure to Interrogate	13/09/2023	13/09/2023
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	13/09/2023	13/09/2023
									A071102	Failure to Interrogate	13/09/2023	13/09/2023
							F12	Serious Injury/ Illness/ Impairment	A070504	Premature Discharge of Battery	13/09/2023	13/09/2023
									A071102	Failure to Interrogate	13/09/2023	13/09/2023
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	13/09/2023	13/09/2023
									A071102	Failure to Interrogate	13/09/2023	13/09/2023
							F23	Unexpected Medical Intervention	A070504	Premature Discharge of Battery	13/09/2023	13/09/2023
									A071102	Failure to Interrogate	13/09/2023	13/09/2023
					CT145	Cardiovascular devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	13/09/2023	13/09/2023
									A071102	Failure to Interrogate	13/09/2023	13/09/2023
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	13/09/2023	13/09/2023
									A071102	Failure to Interrogate	13/09/2023	13/09/2023
							F12	Serious Injury/ Illness/ Impairment	A070504	Premature Discharge of Battery	13/09/2023	13/09/2023
									A071102	Failure to Interrogate	13/09/2023	13/09/2023
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	13/09/2023	13/09/2023
									A071102	Failure to Interrogate	13/09/2023	13/09/2023
							F23	Unexpected Medical Intervention	A070504	Premature Discharge of Battery	13/09/2023	13/09/2023
									A071102	Failure to Interrogate	13/09/2023	13/09/2023
					CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	13/09/2023	13/09/2023
									A071102	Failure to Interrogate	13/09/2023	13/09/2023
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	13/09/2023	13/09/2023
									A071102	Failure to Interrogate	13/09/2023	13/09/2023
							F12	Serious Injury/ Illness/ Impairment	A070504	Premature Discharge of Battery	13/09/2023	13/09/2023
									A071102	Failure to Interrogate	13/09/2023	13/09/2023
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	13/09/2023	13/09/2023
									A071102	Failure to Interrogate	13/09/2023	13/09/2023
							F23	Unexpected Medical Intervention	A070504	Premature Discharge of Battery	13/09/2023	13/09/2023
									A071102	Failure to Interrogate	13/09/2023	13/09/2023
					CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	13/09/2023	13/09/2023
									A071102	Failure to Interrogate	13/09/2023	13/09/2023
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	13/09/2023	13/09/2023
									A071102	Failure to Interrogate	13/09/2023	13/09/2023
							F12	Serious Injury/ Illness/ Impairment	A070504	Premature Discharge of Battery	13/09/2023	13/09/2023
									A071102	Failure to Interrogate	13/09/2023	13/09/2023
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	13/09/2023	13/09/2023
									A071102	Failure to Interrogate	13/09/2023	13/09/2023
							F23	Unexpected Medical Intervention	A070504	Premature Discharge of Battery	13/09/2023	13/09/2023
									A071102	Failure to Interrogate	13/09/2023	13/09/2023
					CT233	Surgical	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	13/09/2023	13/09/2023
									A071102	Failure to Interrogate	13/09/2023	13/09/2023
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	13/09/2023	13/09/2023
									A071102	Failure to Interrogate	13/09/2023	13/09/2023
							F12	Serious Injury/ Illness/ Impairment	A070504	Premature Discharge of Battery	13/09/2023	13/09/2023
									A071102	Failure to Interrogate	13/09/2023	13/09/2023
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	13/09/2023	13/09/2023
									A071102	Failure to Interrogate	13/09/2023	13/09/2023
							F23	Unexpected Medical Intervention	A070504	Premature Discharge of Battery	13/09/2023	13/09/2023
									A071102	Failure to Interrogate	13/09/2023	13/09/2023
					CT2406	Implantable	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	13/09/2023	13/09/2023

								A071102	Failure to Interrogate	13/09/2023	13/09/2023
						F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	13/09/2023	13/09/2023
								A071102	Failure to Interrogate	13/09/2023	13/09/2023
						F12	Serious Injury/ Illness/ Impairment	A070504	Premature Discharge of Battery	13/09/2023	13/09/2023
								A071102	Failure to Interrogate	13/09/2023	13/09/2023
						F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	13/09/2023	13/09/2023
								A071102	Failure to Interrogate	13/09/2023	13/09/2023
						F23	Unexpected Medical Intervention	A070504	Premature Discharge of Battery	13/09/2023	13/09/2023
								A071102	Failure to Interrogate	13/09/2023	13/09/2023
					CT2407	Active implantable	E2403	A070504	Premature Discharge of Battery	13/09/2023	13/09/2023
								A071102	Failure to Interrogate	13/09/2023	13/09/2023
						F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	13/09/2023	13/09/2023
								A071102	Failure to Interrogate	13/09/2023	13/09/2023
						F12	Serious Injury/ Illness/ Impairment	A070504	Premature Discharge of Battery	13/09/2023	13/09/2023
								A071102	Failure to Interrogate	13/09/2023	13/09/2023
						F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	13/09/2023	13/09/2023
								A071102	Failure to Interrogate	13/09/2023	13/09/2023
						F23	Unexpected Medical Intervention	A070504	Premature Discharge of Battery	13/09/2023	13/09/2023
								A071102	Failure to Interrogate	13/09/2023	13/09/2023
					CT257	Implantable cardiac pacemakers	E2403	A070504	Premature Discharge of Battery	13/09/2023	13/09/2023
								A071102	Failure to Interrogate	13/09/2023	13/09/2023
						F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	13/09/2023	13/09/2023
								A071102	Failure to Interrogate	13/09/2023	13/09/2023
						F12	Serious Injury/ Illness/ Impairment	A070504	Premature Discharge of Battery	13/09/2023	13/09/2023
								A071102	Failure to Interrogate	13/09/2023	13/09/2023
						F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	13/09/2023	13/09/2023
								A071102	Failure to Interrogate	13/09/2023	13/09/2023
						F23	Unexpected Medical Intervention	A070504	Premature Discharge of Battery	13/09/2023	13/09/2023
								A071102	Failure to Interrogate	13/09/2023	13/09/2023
					CT2727	Active implant devices/ancillaries	E2403	A070504	Premature Discharge of Battery	13/09/2023	13/09/2023
								A071102	Failure to Interrogate	13/09/2023	13/09/2023
						F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	13/09/2023	13/09/2023
								A071102	Failure to Interrogate	13/09/2023	13/09/2023
						F12	Serious Injury/ Illness/ Impairment	A070504	Premature Discharge of Battery	13/09/2023	13/09/2023
								A071102	Failure to Interrogate	13/09/2023	13/09/2023
						F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	13/09/2023	13/09/2023
								A071102	Failure to Interrogate	13/09/2023	13/09/2023
						F23	Unexpected Medical Intervention	A070504	Premature Discharge of Battery	13/09/2023	13/09/2023
								A071102	Failure to Interrogate	13/09/2023	13/09/2023
					CT2750	Implantable cardiac electrical stimulators and associated devices	E2403	A070504	Premature Discharge of Battery	13/09/2023	13/09/2023
								A071102	Failure to Interrogate	13/09/2023	13/09/2023
						F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	13/09/2023	13/09/2023
								A071102	Failure to Interrogate	13/09/2023	13/09/2023
						F12	Serious Injury/ Illness/ Impairment	A070504	Premature Discharge of Battery	13/09/2023	13/09/2023
								A071102	Failure to Interrogate	13/09/2023	13/09/2023
						F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	13/09/2023	13/09/2023
								A071102	Failure to Interrogate	13/09/2023	13/09/2023
						F23	Unexpected Medical Intervention	A070504	Premature Discharge of Battery	13/09/2023	13/09/2023
								A071102	Failure to Interrogate	13/09/2023	13/09/2023
					CT2866	Implantable cardiac electrical stimulators	E2403	A070504	Premature Discharge of Battery	13/09/2023	13/09/2023
								A071102	Failure to Interrogate	13/09/2023	13/09/2023
						F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	13/09/2023	13/09/2023
								A071102	Failure to Interrogate	13/09/2023	13/09/2023
						F12	Serious Injury/ Illness/ Impairment	A070504	Premature Discharge of Battery	13/09/2023	13/09/2023
								A071102	Failure to Interrogate	13/09/2023	13/09/2023
						F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	13/09/2023	13/09/2023
								A071102	Failure to Interrogate	13/09/2023	13/09/2023
						F23	Unexpected Medical Intervention	A070504	Premature Discharge of Battery	13/09/2023	13/09/2023
								A071102	Failure to Interrogate	13/09/2023	13/09/2023
					CT2985	Active	E2403	A070504	Premature Discharge of Battery	13/09/2023	13/09/2023
								A071102	Failure to Interrogate	13/09/2023	13/09/2023
						F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	13/09/2023	13/09/2023
								A071102	Failure to Interrogate	13/09/2023	13/09/2023
						F12	Serious Injury/ Illness/ Impairment	A070504	Premature Discharge of Battery	13/09/2023	13/09/2023
								A071102	Failure to Interrogate	13/09/2023	13/09/2023
						F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	13/09/2023	13/09/2023
								A071102	Failure to Interrogate	13/09/2023	13/09/2023
						F23	Unexpected Medical Intervention	A070504	Premature Discharge of Battery	13/09/2023	13/09/2023
								A071102	Failure to Interrogate	13/09/2023	13/09/2023
					CT3001	Implantable metal	E2403	A070504	Premature Discharge of Battery	13/09/2023	13/09/2023
								A071102	Failure to Interrogate	13/09/2023	13/09/2023
						F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	13/09/2023	13/09/2023
								A071102	Failure to Interrogate	13/09/2023	13/09/2023
						F12	Serious Injury/ Illness/ Impairment	A070504	Premature Discharge of Battery	13/09/2023	13/09/2023
								A071102	Failure to Interrogate	13/09/2023	13/09/2023
						F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	13/09/2023	13/09/2023
								A071102	Failure to Interrogate	13/09/2023	13/09/2023
						F23	Unexpected Medical Intervention	A070504	Premature Discharge of Battery	13/09/2023	13/09/2023
								A071102	Failure to Interrogate	13/09/2023	13/09/2023
					CT3002	Implantable synthetic polymer	E2403	A070504	Premature Discharge of Battery	13/09/2023	13/09/2023
								A071102	Failure to Interrogate	13/09/2023	13/09/2023
						F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	13/09/2023	13/09/2023
								A071102	Failure to Interrogate	13/09/2023	13/09/2023
						F12	Serious Injury/ Illness/ Impairment	A070504	Premature Discharge of Battery	13/09/2023	13/09/2023
								A071102	Failure to Interrogate	13/09/2023	13/09/2023
						F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	13/09/2023	13/09/2023
								A071102	Failure to Interrogate	13/09/2023	13/09/2023
						F23	Unexpected Medical Intervention	A070504	Premature Discharge of Battery	13/09/2023	13/09/2023

									A071102	Failure to Interrogate	13/09/2023	13/09/2023
					CT3006	Implantable inorganic material	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	13/09/2023	13/09/2023
									A071102	Failure to Interrogate	13/09/2023	13/09/2023
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	13/09/2023	13/09/2023
									A071102	Failure to Interrogate	13/09/2023	13/09/2023
							F12	Serious Injury/ Illness/ Impairment	A070504	Premature Discharge of Battery	13/09/2023	13/09/2023
									A071102	Failure to Interrogate	13/09/2023	13/09/2023
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	13/09/2023	13/09/2023
									A071102	Failure to Interrogate	13/09/2023	13/09/2023
							F23	Unexpected Medical Intervention	A070504	Premature Discharge of Battery	13/09/2023	13/09/2023
									A071102	Failure to Interrogate	13/09/2023	13/09/2023
					CT314	Electrically powered	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	13/09/2023	13/09/2023
									A071102	Failure to Interrogate	13/09/2023	13/09/2023
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	13/09/2023	13/09/2023
									A071102	Failure to Interrogate	13/09/2023	13/09/2023
							F12	Serious Injury/ Illness/ Impairment	A070504	Premature Discharge of Battery	13/09/2023	13/09/2023
									A071102	Failure to Interrogate	13/09/2023	13/09/2023
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	13/09/2023	13/09/2023
									A071102	Failure to Interrogate	13/09/2023	13/09/2023
							F23	Unexpected Medical Intervention	A070504	Premature Discharge of Battery	13/09/2023	13/09/2023
									A071102	Failure to Interrogate	13/09/2023	13/09/2023
					CT321	Long-term surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	13/09/2023	13/09/2023
									A071102	Failure to Interrogate	13/09/2023	13/09/2023
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	13/09/2023	13/09/2023
									A071102	Failure to Interrogate	13/09/2023	13/09/2023
							F12	Serious Injury/ Illness/ Impairment	A070504	Premature Discharge of Battery	13/09/2023	13/09/2023
									A071102	Failure to Interrogate	13/09/2023	13/09/2023
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	13/09/2023	13/09/2023
									A071102	Failure to Interrogate	13/09/2023	13/09/2023
							F23	Unexpected Medical Intervention	A070504	Premature Discharge of Battery	13/09/2023	13/09/2023
									A071102	Failure to Interrogate	13/09/2023	13/09/2023
					CT449	Generators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	13/09/2023	13/09/2023
									A071102	Failure to Interrogate	13/09/2023	13/09/2023
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	13/09/2023	13/09/2023
									A071102	Failure to Interrogate	13/09/2023	13/09/2023
							F12	Serious Injury/ Illness/ Impairment	A070504	Premature Discharge of Battery	13/09/2023	13/09/2023
									A071102	Failure to Interrogate	13/09/2023	13/09/2023
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	13/09/2023	13/09/2023
									A071102	Failure to Interrogate	13/09/2023	13/09/2023
							F23	Unexpected Medical Intervention	A070504	Premature Discharge of Battery	13/09/2023	13/09/2023
									A071102	Failure to Interrogate	13/09/2023	13/09/2023
					CT981	Single use	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	13/09/2023	13/09/2023
									A071102	Failure to Interrogate	13/09/2023	13/09/2023
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	13/09/2023	13/09/2023
									A071102	Failure to Interrogate	13/09/2023	13/09/2023
							F12	Serious Injury/ Illness/ Impairment	A070504	Premature Discharge of Battery	13/09/2023	13/09/2023
									A071102	Failure to Interrogate	13/09/2023	13/09/2023
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	13/09/2023	13/09/2023
									A071102	Failure to Interrogate	13/09/2023	13/09/2023
							F23	Unexpected Medical Intervention	A070504	Premature Discharge of Battery	13/09/2023	13/09/2023
									A071102	Failure to Interrogate	13/09/2023	13/09/2023
					CT983	Surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	13/09/2023	13/09/2023
									A071102	Failure to Interrogate	13/09/2023	13/09/2023
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	13/09/2023	13/09/2023
									A071102	Failure to Interrogate	13/09/2023	13/09/2023
							F12	Serious Injury/ Illness/ Impairment	A070504	Premature Discharge of Battery	13/09/2023	13/09/2023
									A071102	Failure to Interrogate	13/09/2023	13/09/2023
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	13/09/2023	13/09/2023
									A071102	Failure to Interrogate	13/09/2023	13/09/2023
							F23	Unexpected Medical Intervention	A070504	Premature Discharge of Battery	13/09/2023	13/09/2023
									A071102	Failure to Interrogate	13/09/2023	13/09/2023
14076	2023	All other reportable incidents	47265	Dual-chamber implantable pacemaker, rate-responsive	CT141	Electrodes/Leads and associated devices	E0112	Dizziness	A0908	Incorrect, Inadequate or Imprecise Result or Readings	14/02/2023	14/02/2023
									A090801	Signal Artifact/Noise	14/02/2023	14/02/2023
							E011903	Syncope/Fainting	A0908	Incorrect, Inadequate or Imprecise Result or Readings	14/02/2023	14/02/2023
									A090801	Signal Artifact/Noise	14/02/2023	14/02/2023
							F08	Hospitalization or Prolonged Hospitalization	A0908	Incorrect, Inadequate or Imprecise Result or Readings	14/02/2023	14/02/2023
									A090801	Signal Artifact/Noise	14/02/2023	14/02/2023
							F1905	Device Revision or Replacement	A0908	Incorrect, Inadequate or Imprecise Result or Readings	14/02/2023	14/02/2023
									A090801	Signal Artifact/Noise	14/02/2023	14/02/2023
					CT145	Cardiovascular devices	E0112	Dizziness	A0908	Incorrect, Inadequate or Imprecise Result or Readings	14/02/2023	14/02/2023
									A090801	Signal Artifact/Noise	14/02/2023	14/02/2023
							E011903	Syncope/Fainting	A0908	Incorrect, Inadequate or Imprecise Result or Readings	14/02/2023	14/02/2023
									A090801	Signal Artifact/Noise	14/02/2023	14/02/2023
							F08	Hospitalization or Prolonged Hospitalization	A0908	Incorrect, Inadequate or Imprecise Result or Readings	14/02/2023	14/02/2023
									A090801	Signal Artifact/Noise	14/02/2023	14/02/2023
							F1905	Device Revision or Replacement	A0908	Incorrect, Inadequate or Imprecise Result or Readings	14/02/2023	14/02/2023
									A090801	Signal Artifact/Noise	14/02/2023	14/02/2023
					CT1684	Cardiac pacemakers and associated devices	E0112	Dizziness	A0908	Incorrect, Inadequate or Imprecise Result or Readings	14/02/2023	14/02/2023
									A090801	Signal Artifact/Noise	14/02/2023	14/02/2023
							E011903	Syncope/Fainting	A0908	Incorrect, Inadequate or Imprecise Result or Readings	14/02/2023	14/02/2023

									A090801	Signal Artifact/Noise	14/02/2023	14/02/2023
							F08	Hospitalization or Prolonged Hospitalization	A0908	Incorrect, Inadequate or Imprecise Result or Readings	14/02/2023	14/02/2023
									A090801	Signal Artifact/Noise	14/02/2023	14/02/2023
							F1905	Device Revision or Replacement	A0908	Incorrect, Inadequate or Imprecise Result or Readings	14/02/2023	14/02/2023
									A090801	Signal Artifact/Noise	14/02/2023	14/02/2023
					CT1685	Cardiac pacemakers	E0112	Dizziness	A0908	Incorrect, Inadequate or Imprecise Result or Readings	14/02/2023	14/02/2023
									A090801	Signal Artifact/Noise	14/02/2023	14/02/2023
							E011903	Syncope/Fainting	A0908	Incorrect, Inadequate or Imprecise Result or Readings	14/02/2023	14/02/2023
									A090801	Signal Artifact/Noise	14/02/2023	14/02/2023
							F08	Hospitalization or Prolonged Hospitalization	A0908	Incorrect, Inadequate or Imprecise Result or Readings	14/02/2023	14/02/2023
									A090801	Signal Artifact/Noise	14/02/2023	14/02/2023
							F1905	Device Revision or Replacement	A0908	Incorrect, Inadequate or Imprecise Result or Readings	14/02/2023	14/02/2023
									A090801	Signal Artifact/Noise	14/02/2023	14/02/2023
					CT233	Surgical	E0112	Dizziness	A0908	Incorrect, Inadequate or Imprecise Result or Readings	14/02/2023	14/02/2023
									A090801	Signal Artifact/Noise	14/02/2023	14/02/2023
							E011903	Syncope/Fainting	A0908	Incorrect, Inadequate or Imprecise Result or Readings	14/02/2023	14/02/2023
									A090801	Signal Artifact/Noise	14/02/2023	14/02/2023
							F08	Hospitalization or Prolonged Hospitalization	A0908	Incorrect, Inadequate or Imprecise Result or Readings	14/02/2023	14/02/2023
									A090801	Signal Artifact/Noise	14/02/2023	14/02/2023
							F1905	Device Revision or Replacement	A0908	Incorrect, Inadequate or Imprecise Result or Readings	14/02/2023	14/02/2023
									A090801	Signal Artifact/Noise	14/02/2023	14/02/2023
					CT2406	Implantable	E0112	Dizziness	A0908	Incorrect, Inadequate or Imprecise Result or Readings	14/02/2023	14/02/2023
									A090801	Signal Artifact/Noise	14/02/2023	14/02/2023
							E011903	Syncope/Fainting	A0908	Incorrect, Inadequate or Imprecise Result or Readings	14/02/2023	14/02/2023
									A090801	Signal Artifact/Noise	14/02/2023	14/02/2023
							F08	Hospitalization or Prolonged Hospitalization	A0908	Incorrect, Inadequate or Imprecise Result or Readings	14/02/2023	14/02/2023
									A090801	Signal Artifact/Noise	14/02/2023	14/02/2023
							F1905	Device Revision or Replacement	A0908	Incorrect, Inadequate or Imprecise Result or Readings	14/02/2023	14/02/2023
									A090801	Signal Artifact/Noise	14/02/2023	14/02/2023
					CT2407	Active implantable	E0112	Dizziness	A0908	Incorrect, Inadequate or Imprecise Result or Readings	14/02/2023	14/02/2023
									A090801	Signal Artifact/Noise	14/02/2023	14/02/2023
							E011903	Syncope/Fainting	A0908	Incorrect, Inadequate or Imprecise Result or Readings	14/02/2023	14/02/2023
									A090801	Signal Artifact/Noise	14/02/2023	14/02/2023
							F08	Hospitalization or Prolonged Hospitalization	A0908	Incorrect, Inadequate or Imprecise Result or Readings	14/02/2023	14/02/2023
									A090801	Signal Artifact/Noise	14/02/2023	14/02/2023
							F1905	Device Revision or Replacement	A0908	Incorrect, Inadequate or Imprecise Result or Readings	14/02/2023	14/02/2023
									A090801	Signal Artifact/Noise	14/02/2023	14/02/2023
					CT257	Implantable cardiac pacemakers	E0112	Dizziness	A0908	Incorrect, Inadequate or Imprecise Result or Readings	14/02/2023	14/02/2023
									A090801	Signal Artifact/Noise	14/02/2023	14/02/2023
							E011903	Syncope/Fainting	A0908	Incorrect, Inadequate or Imprecise Result or Readings	14/02/2023	14/02/2023
									A090801	Signal Artifact/Noise	14/02/2023	14/02/2023
							F08	Hospitalization or Prolonged Hospitalization	A0908	Incorrect, Inadequate or Imprecise Result or Readings	14/02/2023	14/02/2023
									A090801	Signal Artifact/Noise	14/02/2023	14/02/2023
							F1905	Device Revision or Replacement	A0908	Incorrect, Inadequate or Imprecise Result or Readings	14/02/2023	14/02/2023
									A090801	Signal Artifact/Noise	14/02/2023	14/02/2023
					CT2727	Active implant devices/ancillaries	E0112	Dizziness	A0908	Incorrect, Inadequate or Imprecise Result or Readings	14/02/2023	14/02/2023
									A090801	Signal Artifact/Noise	14/02/2023	14/02/2023
							E011903	Syncope/Fainting	A0908	Incorrect, Inadequate or Imprecise Result or Readings	14/02/2023	14/02/2023
									A090801	Signal Artifact/Noise	14/02/2023	14/02/2023
							F08	Hospitalization or Prolonged Hospitalization	A0908	Incorrect, Inadequate or Imprecise Result or Readings	14/02/2023	14/02/2023
									A090801	Signal Artifact/Noise	14/02/2023	14/02/2023
							F1905	Device Revision or Replacement	A0908	Incorrect, Inadequate or Imprecise Result or Readings	14/02/2023	14/02/2023
									A090801	Signal Artifact/Noise	14/02/2023	14/02/2023
					CT2750	Implantable cardiac electrical stimulators and associated devices	E0112	Dizziness	A0908	Incorrect, Inadequate or Imprecise Result or Readings	14/02/2023	14/02/2023
									A090801	Signal Artifact/Noise	14/02/2023	14/02/2023
							E011903	Syncope/Fainting	A0908	Incorrect, Inadequate or Imprecise Result or Readings	14/02/2023	14/02/2023
									A090801	Signal Artifact/Noise	14/02/2023	14/02/2023
							F08	Hospitalization or Prolonged Hospitalization	A0908	Incorrect, Inadequate or Imprecise Result or Readings	14/02/2023	14/02/2023
									A090801	Signal Artifact/Noise	14/02/2023	14/02/2023
							F1905	Device Revision or Replacement	A0908	Incorrect, Inadequate or Imprecise Result or Readings	14/02/2023	14/02/2023
									A090801	Signal Artifact/Noise	14/02/2023	14/02/2023
					CT2866	Implantable cardiac electrical stimulators	E0112	Dizziness	A0908	Incorrect, Inadequate or Imprecise Result or Readings	14/02/2023	14/02/2023
									A090801	Signal Artifact/Noise	14/02/2023	14/02/2023

							E011903	Syncope/Fainting	A0908	Incorrect, Inadequate or Imprecise Result or Readings	14/02/2023	14/02/2023
									A090801	Signal Artifact/Noise	14/02/2023	14/02/2023
							F08	Hospitalization or Prolonged Hospitalization	A0908	Incorrect, Inadequate or Imprecise Result or Readings	14/02/2023	14/02/2023
									A090801	Signal Artifact/Noise	14/02/2023	14/02/2023
							F1905	Device Revision or Replacement	A0908	Incorrect, Inadequate or Imprecise Result or Readings	14/02/2023	14/02/2023
									A090801	Signal Artifact/Noise	14/02/2023	14/02/2023
						CT2985	Active	E0112	Dizziness	A0908	Incorrect, Inadequate or Imprecise Result or Readings	14/02/2023	14/02/2023
										A090801	Signal Artifact/Noise	14/02/2023	14/02/2023
								E011903	Syncope/Fainting	A0908	Incorrect, Inadequate or Imprecise Result or Readings	14/02/2023	14/02/2023
										A090801	Signal Artifact/Noise	14/02/2023	14/02/2023
								F08	Hospitalization or Prolonged Hospitalization	A0908	Incorrect, Inadequate or Imprecise Result or Readings	14/02/2023	14/02/2023
										A090801	Signal Artifact/Noise	14/02/2023	14/02/2023
								F1905	Device Revision or Replacement	A0908	Incorrect, Inadequate or Imprecise Result or Readings	14/02/2023	14/02/2023
										A090801	Signal Artifact/Noise	14/02/2023	14/02/2023
						CT3001	Implantable metal	E0112	Dizziness	A0908	Incorrect, Inadequate or Imprecise Result or Readings	14/02/2023	14/02/2023
										A090801	Signal Artifact/Noise	14/02/2023	14/02/2023
								E011903	Syncope/Fainting	A0908	Incorrect, Inadequate or Imprecise Result or Readings	14/02/2023	14/02/2023
										A090801	Signal Artifact/Noise	14/02/2023	14/02/2023
								F08	Hospitalization or Prolonged Hospitalization	A0908	Incorrect, Inadequate or Imprecise Result or Readings	14/02/2023	14/02/2023
										A090801	Signal Artifact/Noise	14/02/2023	14/02/2023
								F1905	Device Revision or Replacement	A0908	Incorrect, Inadequate or Imprecise Result or Readings	14/02/2023	14/02/2023
										A090801	Signal Artifact/Noise	14/02/2023	14/02/2023
						CT3002	Implantable synthetic polymer	E0112	Dizziness	A0908	Incorrect, Inadequate or Imprecise Result or Readings	14/02/2023	14/02/2023
										A090801	Signal Artifact/Noise	14/02/2023	14/02/2023
								E011903	Syncope/Fainting	A0908	Incorrect, Inadequate or Imprecise Result or Readings	14/02/2023	14/02/2023
										A090801	Signal Artifact/Noise	14/02/2023	14/02/2023
								F08	Hospitalization or Prolonged Hospitalization	A0908	Incorrect, Inadequate or Imprecise Result or Readings	14/02/2023	14/02/2023
										A090801	Signal Artifact/Noise	14/02/2023	14/02/2023
								F1905	Device Revision or Replacement	A0908	Incorrect, Inadequate or Imprecise Result or Readings	14/02/2023	14/02/2023
										A090801	Signal Artifact/Noise	14/02/2023	14/02/2023
						CT3006	Implantable inorganic material	E0112	Dizziness	A0908	Incorrect, Inadequate or Imprecise Result or Readings	14/02/2023	14/02/2023
										A090801	Signal Artifact/Noise	14/02/2023	14/02/2023
								E011903	Syncope/Fainting	A0908	Incorrect, Inadequate or Imprecise Result or Readings	14/02/2023	14/02/2023
										A090801	Signal Artifact/Noise	14/02/2023	14/02/2023
								F08	Hospitalization or Prolonged Hospitalization	A0908	Incorrect, Inadequate or Imprecise Result or Readings	14/02/2023	14/02/2023
										A090801	Signal Artifact/Noise	14/02/2023	14/02/2023
								F1905	Device Revision or Replacement	A0908	Incorrect, Inadequate or Imprecise Result or Readings	14/02/2023	14/02/2023
										A090801	Signal Artifact/Noise	14/02/2023	14/02/2023
						CT314	Electrically powered	E0112	Dizziness	A0908	Incorrect, Inadequate or Imprecise Result or Readings	14/02/2023	14/02/2023
										A090801	Signal Artifact/Noise	14/02/2023	14/02/2023
								E011903	Syncope/Fainting	A0908	Incorrect, Inadequate or Imprecise Result or Readings	14/02/2023	14/02/2023
										A090801	Signal Artifact/Noise	14/02/2023	14/02/2023
								F08	Hospitalization or Prolonged Hospitalization	A0908	Incorrect, Inadequate or Imprecise Result or Readings	14/02/2023	14/02/2023
										A090801	Signal Artifact/Noise	14/02/2023	14/02/2023
								F1905	Device Revision or Replacement	A0908	Incorrect, Inadequate or Imprecise Result or Readings	14/02/2023	14/02/2023
										A090801	Signal Artifact/Noise	14/02/2023	14/02/2023
						CT321	Long-term surgical invasive	E0112	Dizziness	A0908	Incorrect, Inadequate or Imprecise Result or Readings	14/02/2023	14/02/2023
										A090801	Signal Artifact/Noise	14/02/2023	14/02/2023
								E011903	Syncope/Fainting	A0908	Incorrect, Inadequate or Imprecise Result or Readings	14/02/2023	14/02/2023
										A090801	Signal Artifact/Noise	14/02/2023	14/02/2023
								F08	Hospitalization or Prolonged Hospitalization	A0908	Incorrect, Inadequate or Imprecise Result or Readings	14/02/2023	14/02/2023
										A090801	Signal Artifact/Noise	14/02/2023	14/02/2023
								F1905	Device Revision or Replacement	A0908	Incorrect, Inadequate or Imprecise Result or Readings	14/02/2023	14/02/2023
										A090801	Signal Artifact/Noise	14/02/2023	14/02/2023
						CT449	Generators and associated devices	E0112	Dizziness	A0908	Incorrect, Inadequate or Imprecise Result or Readings	14/02/2023	14/02/2023
										A090801	Signal Artifact/Noise	14/02/2023	14/02/2023
								E011903	Syncope/Fainting	A0908	Incorrect, Inadequate or Imprecise Result or Readings	14/02/2023	14/02/2023
										A090801	Signal Artifact/Noise	14/02/2023	14/02/2023
								F08	Hospitalization or Prolonged Hospitalization	A0908	Incorrect, Inadequate or Imprecise Result or Readings	14/02/2023	14/02/2023
										A090801	Signal Artifact/Noise	14/02/2023	14/02/2023
								F1905	Device Revision or Replacement	A0908	Incorrect, Inadequate or Imprecise Result or Readings	14/02/2023	14/02/2023
										A090801	Signal Artifact/Noise	14/02/2023	14/02/2023
						CT981	Single use	E0112	Dizziness	A0908	Incorrect, Inadequate or Imprecise Result or Readings	14/02/2023	14/02/2023

								A090801	Signal Artifact/Noise	14/02/2023	14/02/2023
						E011903	Syncope/Fainting	A0908	Incorrect, Inadequate or Imprecise Result or Readings	14/02/2023	14/02/2023
								A090801	Signal Artifact/Noise	14/02/2023	14/02/2023
						F08	Hospitalization or Prolonged Hospitalization	A0908	Incorrect, Inadequate or Imprecise Result or Readings	14/02/2023	14/02/2023
								A090801	Signal Artifact/Noise	14/02/2023	14/02/2023
						F1905	Device Revision or Replacement	A0908	Incorrect, Inadequate or Imprecise Result or Readings	14/02/2023	14/02/2023
								A090801	Signal Artifact/Noise	14/02/2023	14/02/2023
				CT983	Surgical invasive	E0112	Dizziness	A0908	Incorrect, Inadequate or Imprecise Result or Readings	14/02/2023	14/02/2023
								A090801	Signal Artifact/Noise	14/02/2023	14/02/2023
						E011903	Syncope/Fainting	A0908	Incorrect, Inadequate or Imprecise Result or Readings	14/02/2023	14/02/2023
								A090801	Signal Artifact/Noise	14/02/2023	14/02/2023
						F08	Hospitalization or Prolonged Hospitalization	A0908	Incorrect, Inadequate or Imprecise Result or Readings	14/02/2023	14/02/2023
								A090801	Signal Artifact/Noise	14/02/2023	14/02/2023
						F1905	Device Revision or Replacement	A0908	Incorrect, Inadequate or Imprecise Result or Readings	14/02/2023	14/02/2023
								A090801	Signal Artifact/Noise	14/02/2023	14/02/2023
14077	2023	All other reportable incidents	36241	Pacemaker/defibrillator lead	CT1684	Cardiac pacemakers and associated devices	E0604	Cardiac Perforation	A051201	Device Dislodged or Dislocated	09/06/2017	09/06/2017
							F08	Hospitalization or Prolonged Hospitalization	A051201	Device Dislodged or Dislocated	09/06/2017	09/06/2017
							F1905	Device Revision or Replacement	A051201	Device Dislodged or Dislocated	09/06/2017	09/06/2017
14078	2023	All other reportable incidents	36241	Pacemaker/defibrillator lead	CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A040101	Fracture		
									A070101	Failure to Capture		
									A070909	Oversensing		
									A0712	Pacing Problem		
									A072201	High Impedance		
									A090801	Signal Artifact/Noise		
							F08	Hospitalization or Prolonged Hospitalization	A040101	Fracture		
									A070101	Failure to Capture		
									A070909	Oversensing		
									A0712	Pacing Problem		
									A072201	High Impedance		
									A090801	Signal Artifact/Noise		
							F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A040101	Fracture		
									A070101	Failure to Capture		
									A070909	Oversensing		
									A0712	Pacing Problem		
									A072201	High Impedance		
									A090801	Signal Artifact/Noise		
							F1905	Device Revision or Replacement	A040101	Fracture		
									A070101	Failure to Capture		
									A070909	Oversensing		
									A0712	Pacing Problem		
									A072201	High Impedance		
									A090801	Signal Artifact/Noise		
							F2203	Imaging Required	A040101	Fracture		
									A070101	Failure to Capture		
									A070909	Oversensing		
									A0712	Pacing Problem		
									A072201	High Impedance		
									A090801	Signal Artifact/Noise		
14079	2023	All other reportable incidents	47265	Dual-chamber implantable pacemaker, rate-responsive	CT1685	Cardiac pacemakers	E2103	Device Overstimulation of Tissue	A0908	Incorrect, Inadequate or Imprecise Result or Readings	04/09/2023	04/09/2023
							F08	Hospitalization or Prolonged Hospitalization	A0908	Incorrect, Inadequate or Imprecise Result or Readings	04/09/2023	04/09/2023
							F1905	Device Revision or Replacement	A0908	Incorrect, Inadequate or Imprecise Result or Readings	04/09/2023	04/09/2023
					CT257	Implantable cardiac pacemakers	E2103	Device Overstimulation of Tissue	A0908	Incorrect, Inadequate or Imprecise Result or Readings	04/09/2023	04/09/2023
							F08	Hospitalization or Prolonged Hospitalization	A0908	Incorrect, Inadequate or Imprecise Result or Readings	04/09/2023	04/09/2023
							F1905	Device Revision or Replacement	A0908	Incorrect, Inadequate or Imprecise Result or Readings	04/09/2023	04/09/2023
					CT2866	Implantable cardiac electrical stimulators	E2103	Device Overstimulation of Tissue	A0908	Incorrect, Inadequate or Imprecise Result or Readings	04/09/2023	04/09/2023
							F08	Hospitalization or Prolonged Hospitalization	A0908	Incorrect, Inadequate or Imprecise Result or Readings	04/09/2023	04/09/2023
							F1905	Device Revision or Replacement	A0908	Incorrect, Inadequate or Imprecise Result or Readings	04/09/2023	04/09/2023
14080	2023	All other reportable incidents	36241	Pacemaker/defibrillator lead	CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A05	Mechanical Problem	12/06/2017	12/06/2017
									A051201	Device Dislodged or Dislocated	12/06/2017	12/06/2017
									A1502	Positioning Problem	12/06/2017	12/06/2017
							F08	Hospitalization or Prolonged Hospitalization	A05	Mechanical Problem	12/06/2017	12/06/2017
									A051201	Device Dislodged or Dislocated	12/06/2017	12/06/2017
									A1502	Positioning Problem	12/06/2017	12/06/2017
							F1905	Device Revision or Replacement	A05	Mechanical Problem	12/06/2017	12/06/2017
									A051201	Device Dislodged or Dislocated	12/06/2017	12/06/2017
									A1502	Positioning Problem	12/06/2017	12/06/2017
14081	2023	All other reportable incidents	47263	Cardiac resynchronization therapy implantable pacemaker	CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A0908	Incorrect, Inadequate or Imprecise Result or Readings	12/10/2022	12/10/2022
							F08	Hospitalization or Prolonged Hospitalization	A0908	Incorrect, Inadequate or Imprecise Result or Readings	12/10/2022	12/10/2022
							F1905	Device Revision or Replacement	A0908	Incorrect, Inadequate or Imprecise Result or Readings	12/10/2022	12/10/2022
					CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A0908	Incorrect, Inadequate or Imprecise Result or Readings	12/10/2022	12/10/2022
							F08	Hospitalization or Prolonged Hospitalization	A0908	Incorrect, Inadequate or Imprecise Result or Readings	12/10/2022	12/10/2022

						F1905	Device Revision or Replacement	A0908	Incorrect, Inadequate or Imprecise Result or Readings	12/10/2022	12/10/2022
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A0908	Incorrect, Inadequate or Imprecise Result or Readings	12/10/2022	12/10/2022
							F08	Hospitalization or Prolonged Hospitalization	A0908	Incorrect, Inadequate or Imprecise Result or Readings	12/10/2022	12/10/2022
							F1905	Device Revision or Replacement	A0908	Incorrect, Inadequate or Imprecise Result or Readings	12/10/2022	12/10/2022
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A0908	Incorrect, Inadequate or Imprecise Result or Readings	12/10/2022	12/10/2022
							F08	Hospitalization or Prolonged Hospitalization	A0908	Incorrect, Inadequate or Imprecise Result or Readings	12/10/2022	12/10/2022
							F1905	Device Revision or Replacement	A0908	Incorrect, Inadequate or Imprecise Result or Readings	12/10/2022	12/10/2022
14082	2023	All other reportable incidents	35223	Endocardial/interventricular septal pacing lead	CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070102	High Capture Threshold		
									A070909	Oversensing		
									A0722	Impedance Problem		
							F26	No Health Consequences or Impact	A070102	High Capture Threshold		
									A070909	Oversensing		
									A0722	Impedance Problem		
14083	2023	All other reportable incidents	17882	Physiologic-monitoring defibrillation system	CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A0719	Unexpected Shutdown		
							F27	Problem Identified During Non-Clinical Procedure	A0719	Unexpected Shutdown		
14084	2023		47265	Dual-chamber implantable pacemaker, rate-responsive	CT141	Electrodes/Leads and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	19/09/2023	19/09/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	19/09/2023	19/09/2023
					CT145	Cardiovascular devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	19/09/2023	19/09/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	19/09/2023	19/09/2023
					CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	19/09/2023	19/09/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	19/09/2023	19/09/2023
					CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	19/09/2023	19/09/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	19/09/2023	19/09/2023
					CT233	Surgical	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	19/09/2023	19/09/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	19/09/2023	19/09/2023
					CT2406	Implantable	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	19/09/2023	19/09/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	19/09/2023	19/09/2023
					CT2407	Active implantable	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	19/09/2023	19/09/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	19/09/2023	19/09/2023
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	19/09/2023	19/09/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	19/09/2023	19/09/2023
					CT2727	Active implant devices/ancillaries	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	19/09/2023	19/09/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	19/09/2023	19/09/2023
					CT2750	Implantable cardiac electrical stimulators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	19/09/2023	19/09/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	19/09/2023	19/09/2023
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	19/09/2023	19/09/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	19/09/2023	19/09/2023
					CT2985	Active	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	19/09/2023	19/09/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	19/09/2023	19/09/2023
					CT3001	Implantable metal	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	19/09/2023	19/09/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	19/09/2023	19/09/2023
					CT3002	Implantable synthetic polymer	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	19/09/2023	19/09/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	19/09/2023	19/09/2023
					CT3006	Implantable inorganic material	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	19/09/2023	19/09/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	19/09/2023	19/09/2023
					CT314	Electrically powered	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	19/09/2023	19/09/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	19/09/2023	19/09/2023
					CT321	Long-term surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	19/09/2023	19/09/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	19/09/2023	19/09/2023
					CT449	Generators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	19/09/2023	19/09/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	19/09/2023	19/09/2023
					CT981	Single use	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	19/09/2023	19/09/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	19/09/2023	19/09/2023
					CT983	Surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	19/09/2023	19/09/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	19/09/2023	19/09/2023
14085	2023		47265	Dual-chamber implantable pacemaker, rate-responsive	CT141	Electrodes/Leads and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	19/09/2023	19/09/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	19/09/2023	19/09/2023
					CT145	Cardiovascular devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	19/09/2023	19/09/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	19/09/2023	19/09/2023
					CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	19/09/2023	19/09/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	19/09/2023	19/09/2023
					CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	19/09/2023	19/09/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	19/09/2023	19/09/2023
					CT233	Surgical	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	19/09/2023	19/09/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	19/09/2023	19/09/2023
					CT2406	Implantable	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	19/09/2023	19/09/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	19/09/2023	19/09/2023
					CT2407	Active implantable	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	19/09/2023	19/09/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	19/09/2023	19/09/2023
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	19/09/2023	19/09/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	19/09/2023	19/09/2023
					CT2727	Active implant devices/ancillaries	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	19/09/2023	19/09/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	19/09/2023	19/09/2023
					CT2750	Implantable cardiac electrical stimulators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	19/09/2023	19/09/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	19/09/2023	19/09/2023
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	19/09/2023	19/09/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	19/09/2023	19/09/2023
					CT2985	Active	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	19/09/2023	19/09/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	19/09/2023	19/09/2023
					CT3001	Implantable metal	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	19/09/2023	19/09/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	19/09/2023	19/09/2023
					CT3002	Implantable synthetic polymer	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	19/09/2023	19/09/2023

						F1901	Additional Surgery	A24	No Apparent Device Problem	19/09/2023	19/09/2023
					CT3006	Implantable inorganic material	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	19/09/2023	19/09/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	19/09/2023	19/09/2023
					CT314	Electrically powered	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	19/09/2023	19/09/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	19/09/2023	19/09/2023
					CT321	Long-term surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	19/09/2023	19/09/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	19/09/2023	19/09/2023
					CT449	Generators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	19/09/2023	19/09/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	19/09/2023	19/09/2023
					CT981	Single use	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	19/09/2023	19/09/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	19/09/2023	19/09/2023
					CT983	Surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	19/09/2023	19/09/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	19/09/2023	19/09/2023
14086	2023		47265	Dual-chamber implantable pacemaker, rate-responsive	CT141	Electrodes/Leads and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	19/09/2023	19/09/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	19/09/2023	19/09/2023
					CT145	Cardiovascular devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	19/09/2023	19/09/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	19/09/2023	19/09/2023
					CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	19/09/2023	19/09/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	19/09/2023	19/09/2023
					CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	19/09/2023	19/09/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	19/09/2023	19/09/2023
					CT233	Surgical	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	19/09/2023	19/09/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	19/09/2023	19/09/2023
					CT2406	Implantable	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	19/09/2023	19/09/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	19/09/2023	19/09/2023
					CT2407	Active implantable	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	19/09/2023	19/09/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	19/09/2023	19/09/2023
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	19/09/2023	19/09/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	19/09/2023	19/09/2023
					CT2727	Active implant devices/ancillaries	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	19/09/2023	19/09/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	19/09/2023	19/09/2023
					CT2750	Implantable cardiac electrical stimulators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	19/09/2023	19/09/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	19/09/2023	19/09/2023
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	19/09/2023	19/09/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	19/09/2023	19/09/2023
					CT2985	Active	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	19/09/2023	19/09/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	19/09/2023	19/09/2023
					CT3001	Implantable metal	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	19/09/2023	19/09/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	19/09/2023	19/09/2023
					CT3002	Implantable synthetic polymer	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	19/09/2023	19/09/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	19/09/2023	19/09/2023
					CT3006	Implantable inorganic material	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	19/09/2023	19/09/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	19/09/2023	19/09/2023
					CT314	Electrically powered	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	19/09/2023	19/09/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	19/09/2023	19/09/2023
					CT321	Long-term surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	19/09/2023	19/09/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	19/09/2023	19/09/2023
					CT449	Generators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	19/09/2023	19/09/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	19/09/2023	19/09/2023
					CT981	Single use	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	19/09/2023	19/09/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	19/09/2023	19/09/2023
					CT983	Surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	19/09/2023	19/09/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	19/09/2023	19/09/2023
14087	2023	All other reportable incidents	63188	Left-ventricular wireless pacing system sensor/transmitter	CT141	Electrodes/Leads and associated devices			A0708	Power Problem		
					CT145	Cardiovascular devices			A0708	Power Problem		
					CT1577	Wireless monitoring system transmitters			A0708	Power Problem		
					CT1684	Cardiac pacemakers and associated devices			A0708	Power Problem		
					CT2317	Transmitters/Receivers/Transceivers			A0708	Power Problem		
					CT233	Surgical			A0708	Power Problem		
					CT2406	Implantable			A0708	Power Problem		
					CT2407	Active implantable			A0708	Power Problem		
					CT255	Transducers and associated devices			A0708	Power Problem		
					CT2726	Physiological monitoring/measuring devices			A0708	Power Problem		
					CT2727	Active implant devices/ancillaries			A0708	Power Problem		
					CT2750	Implantable cardiac electrical stimulators and associated devices			A0708	Power Problem		
					CT2885	Physiological monitoring/measuring system transmitters/receivers/transceivers			A0708	Power Problem		
					CT2985	Active			A0708	Power Problem		
					CT3001	Implantable metal			A0708	Power Problem		
					CT3002	Implantable synthetic polymer			A0708	Power Problem		
					CT3006	Implantable inorganic material			A0708	Power Problem		
					CT314	Electrically powered			A0708	Power Problem		
					CT321	Long-term surgical invasive			A0708	Power Problem		
					CT384	Sensors and associated devices			A0708	Power Problem		
					CT449	Generators and associated devices			A0708	Power Problem		
					CT981	Single use			A0708	Power Problem		
					CT983	Surgical invasive			A0708	Power Problem		
					CT994	Sensors			A0708	Power Problem		
14088	2023	All other reportable incidents	35852	Single-chamber implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
14089	2023	All other reportable incidents	35852	Single-chamber implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A0722	Impedance Problem		
									A072201	High Impedance		
							F26	No Health Consequences or Impact	A0722	Impedance Problem		
									A072201	High Impedance		
14090	2023	All other reportable incidents	35224	External pacemaker, epicardial pacing	CT1685	Cardiac pacemakers	E0602	Cardiac Arrest	A0709	Device Sensing Problem		

									A0902	Display or Visual Feedback Problem		
							F1203	Life Threatening Illness or Injury	A0709	Device Sensing Problem		
									A0902	Display or Visual Feedback Problem		
							F23	Unexpected Medical Intervention	A0709	Device Sensing Problem		
									A0902	Display or Visual Feedback Problem		
14091	2023	Death	17882	Physiologic-monitoring defibrillation system	CT1684	Cardiac pacemakers and associated devices			A071204	No Pacing		
14092	2023	All other reportable incidents	36241	Pacemaker/defibrillator lead	CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A072201	High Impedance		
							F26	No Health Consequences or Impact	A072201	High Impedance		
14093	2023	All other reportable incidents	35223	Endocardial/interventricular septal pacing lead	CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A0401	Break		
									A070909	Oversensing		
							F26	No Health Consequences or Impact	A0401	Break		
									A070909	Oversensing		
14094	2023	All other reportable incidents	35223	Endocardial/interventricular septal pacing lead	CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070909	Oversensing		
							F26	No Health Consequences or Impact	A070909	Oversensing		
14095	2023	All other reportable incidents	35852	Single-chamber implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	30/06/2023	30/06/2023
									A0908	Incorrect, Inadequate or Imprecise Result or Readings	30/06/2023	30/06/2023
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	30/06/2023	30/06/2023
									A0908	Incorrect, Inadequate or Imprecise Result or Readings	30/06/2023	30/06/2023
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	30/06/2023	30/06/2023
									A0908	Incorrect, Inadequate or Imprecise Result or Readings	30/06/2023	30/06/2023
14096	2023	All other reportable incidents	47265	Dual-chamber implantable pacemaker, rate-responsive	CT141	Electrodes/Leads and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070101	Failure to Capture		
									A070102	High Capture Threshold		
							F26	No Health Consequences or Impact	A070101	Failure to Capture		
									A070102	High Capture Threshold		
					CT145	Cardiovascular devices	E2403	No Clinical Signs, Symptoms or Conditions	A070101	Failure to Capture		
									A070102	High Capture Threshold		
							F26	No Health Consequences or Impact	A070101	Failure to Capture		
									A070102	High Capture Threshold		
					CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070101	Failure to Capture		
									A070102	High Capture Threshold		
							F26	No Health Consequences or Impact	A070101	Failure to Capture		
									A070102	High Capture Threshold		
					CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A070101	Failure to Capture		
									A070102	High Capture Threshold		
							F26	No Health Consequences or Impact	A070101	Failure to Capture		
									A070102	High Capture Threshold		
					CT233	Surgical	E2403	No Clinical Signs, Symptoms or Conditions	A070101	Failure to Capture		
									A070102	High Capture Threshold		
							F26	No Health Consequences or Impact	A070101	Failure to Capture		
									A070102	High Capture Threshold		
					CT2406	Implantable	E2403	No Clinical Signs, Symptoms or Conditions	A070101	Failure to Capture		
									A070102	High Capture Threshold		
							F26	No Health Consequences or Impact	A070101	Failure to Capture		
									A070102	High Capture Threshold		
					CT2407	Active implantable	E2403	No Clinical Signs, Symptoms or Conditions	A070101	Failure to Capture		
									A070102	High Capture Threshold		
							F26	No Health Consequences or Impact	A070101	Failure to Capture		
									A070102	High Capture Threshold		
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A070101	Failure to Capture		
									A070102	High Capture Threshold		
							F26	No Health Consequences or Impact	A070101	Failure to Capture		
									A070102	High Capture Threshold		
					CT2727	Active implant devices/ancillaries	E2403	No Clinical Signs, Symptoms or Conditions	A070101	Failure to Capture		
									A070102	High Capture Threshold		
							F26	No Health Consequences or Impact	A070101	Failure to Capture		
									A070102	High Capture Threshold		
					CT2750	Implantable cardiac electrical stimulators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070101	Failure to Capture		
									A070102	High Capture Threshold		
							F26	No Health Consequences or Impact	A070101	Failure to Capture		
									A070102	High Capture Threshold		
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A070101	Failure to Capture		
									A070102	High Capture Threshold		
							F26	No Health Consequences or Impact	A070101	Failure to Capture		
									A070102	High Capture Threshold		
					CT2985	Active	E2403	No Clinical Signs, Symptoms or Conditions	A070101	Failure to Capture		
									A070102	High Capture Threshold		
							F26	No Health Consequences or Impact	A070101	Failure to Capture		
									A070102	High Capture Threshold		
					CT3001	Implantable metal	E2403	No Clinical Signs, Symptoms or Conditions	A070101	Failure to Capture		
									A070102	High Capture Threshold		
							F26	No Health Consequences or Impact	A070101	Failure to Capture		
									A070102	High Capture Threshold		
					CT3002	Implantable synthetic polymer	E2403	No Clinical Signs, Symptoms or Conditions	A070101	Failure to Capture		
									A070102	High Capture Threshold		
							F26	No Health Consequences or Impact	A070101	Failure to Capture		
									A070102	High Capture Threshold		
					CT3006	Implantable inorganic material	E2403	No Clinical Signs, Symptoms or Conditions	A070101	Failure to Capture		
									A070102	High Capture Threshold		
							F26	No Health Consequences or Impact	A070101	Failure to Capture		
									A070102	High Capture Threshold		
					CT314	Electrically powered	E2403	No Clinical Signs, Symptoms or Conditions	A070101	Failure to Capture		
									A070102	High Capture Threshold		
							F26	No Health Consequences or Impact	A070101	Failure to Capture		
									A070102	High Capture Threshold		
					CT321	Long-term surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A070101	Failure to Capture		

								A070102	High Capture Threshold		
						F26	No Health Consequences or Impact	A070101	Failure to Capture		
								A070102	High Capture Threshold		
					CT449	Generators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070101	Failure to Capture		
									A070102	High Capture Threshold		
							F26	No Health Consequences or Impact	A070101	Failure to Capture		
									A070102	High Capture Threshold		
					CT981	Single use	E2403	No Clinical Signs, Symptoms or Conditions	A070101	Failure to Capture		
									A070102	High Capture Threshold		
							F26	No Health Consequences or Impact	A070101	Failure to Capture		
									A070102	High Capture Threshold		
					CT983	Surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A070101	Failure to Capture		
									A070102	High Capture Threshold		
							F26	No Health Consequences or Impact	A070101	Failure to Capture		
									A070102	High Capture Threshold		
14097	2023	All other reportable incidents	36241	Pacemaker/defibrillator lead	CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070101	Failure to Capture		
									A070102	High Capture Threshold		
							F26	No Health Consequences or Impact	A070101	Failure to Capture		
									A070102	High Capture Threshold		
14098	2023	All other reportable incidents	35852	Single-chamber implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A0908	Incorrect, Inadequate or Imprecise Result or Readings	14/09/2023	14/09/2023
							F08	Hospitalization or Prolonged Hospitalization	A0908	Incorrect, Inadequate or Imprecise Result or Readings	14/09/2023	14/09/2023
							F1903	Device Explantation	A0908	Incorrect, Inadequate or Imprecise Result or Readings	14/09/2023	14/09/2023
14099	2023	All other reportable incidents	35223	Endocardial/interventricular septal pacing lead	CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070909	Oversensing		
									A070910	Undersensing		
							F12	Serious Injury/ Illness/ Impairment	A070909	Oversensing		
									A070910	Undersensing		
							F23	Unexpected Medical Intervention	A070909	Oversensing		
									A070910	Undersensing		
14100	2023	All other reportable incidents	47265	Dual-chamber implantable pacemaker, rate-responsive	CT141	Electrodes/Leads and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	07/08/2023	07/08/2023
									A071102	Failure to Interrogate	07/08/2023	07/08/2023
							F1901	Additional Surgery	A070504	Premature Discharge of Battery	07/08/2023	07/08/2023
									A071102	Failure to Interrogate	07/08/2023	07/08/2023
					CT145	Cardiovascular devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	07/08/2023	07/08/2023
									A071102	Failure to Interrogate	07/08/2023	07/08/2023
							F1901	Additional Surgery	A070504	Premature Discharge of Battery	07/08/2023	07/08/2023
									A071102	Failure to Interrogate	07/08/2023	07/08/2023
					CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	07/08/2023	07/08/2023
									A071102	Failure to Interrogate	07/08/2023	07/08/2023
							F1901	Additional Surgery	A070504	Premature Discharge of Battery	07/08/2023	07/08/2023
									A071102	Failure to Interrogate	07/08/2023	07/08/2023
					CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	07/08/2023	07/08/2023
									A071102	Failure to Interrogate	07/08/2023	07/08/2023
							F1901	Additional Surgery	A070504	Premature Discharge of Battery	07/08/2023	07/08/2023
									A071102	Failure to Interrogate	07/08/2023	07/08/2023
					CT233	Surgical	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	07/08/2023	07/08/2023
									A071102	Failure to Interrogate	07/08/2023	07/08/2023
							F1901	Additional Surgery	A070504	Premature Discharge of Battery	07/08/2023	07/08/2023
									A071102	Failure to Interrogate	07/08/2023	07/08/2023
					CT2406	Implantable	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	07/08/2023	07/08/2023
									A071102	Failure to Interrogate	07/08/2023	07/08/2023
							F1901	Additional Surgery	A070504	Premature Discharge of Battery	07/08/2023	07/08/2023
									A071102	Failure to Interrogate	07/08/2023	07/08/2023
					CT2407	Active implantable	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	07/08/2023	07/08/2023
									A071102	Failure to Interrogate	07/08/2023	07/08/2023
							F1901	Additional Surgery	A070504	Premature Discharge of Battery	07/08/2023	07/08/2023
									A071102	Failure to Interrogate	07/08/2023	07/08/2023
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	07/08/2023	07/08/2023
									A071102	Failure to Interrogate	07/08/2023	07/08/2023
							F1901	Additional Surgery	A070504	Premature Discharge of Battery	07/08/2023	07/08/2023
									A071102	Failure to Interrogate	07/08/2023	07/08/2023
					CT2727	Active implant devices/ancillaries	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	07/08/2023	07/08/2023
									A071102	Failure to Interrogate	07/08/2023	07/08/2023
							F1901	Additional Surgery	A070504	Premature Discharge of Battery	07/08/2023	07/08/2023
									A071102	Failure to Interrogate	07/08/2023	07/08/2023
					CT2750	Implantable cardiac electrical stimulators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	07/08/2023	07/08/2023
									A071102	Failure to Interrogate	07/08/2023	07/08/2023
							F1901	Additional Surgery	A070504	Premature Discharge of Battery	07/08/2023	07/08/2023
									A071102	Failure to Interrogate	07/08/2023	07/08/2023
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	07/08/2023	07/08/2023
									A071102	Failure to Interrogate	07/08/2023	07/08/2023
							F1901	Additional Surgery	A070504	Premature Discharge of Battery	07/08/2023	07/08/2023
									A071102	Failure to Interrogate	07/08/2023	07/08/2023
					CT2985	Active	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	07/08/2023	07/08/2023
									A071102	Failure to Interrogate	07/08/2023	07/08/2023
							F1901	Additional Surgery	A070504	Premature Discharge of Battery	07/08/2023	07/08/2023
									A071102	Failure to Interrogate	07/08/2023	07/08/2023
					CT3001	Implantable metal	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	07/08/2023	07/08/2023
									A071102	Failure to Interrogate	07/08/2023	07/08/2023
							F1901	Additional Surgery	A070504	Premature Discharge of Battery	07/08/2023	07/08/2023
									A071102	Failure to Interrogate	07/08/2023	07/08/2023
					CT3002	Implantable synthetic polymer	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	07/08/2023	07/08/2023
									A071102	Failure to Interrogate	07/08/2023	07/08/2023
							F1901	Additional Surgery	A070504	Premature Discharge of Battery	07/08/2023	07/08/2023
									A071102	Failure to Interrogate	07/08/2023	07/08/2023
					CT3006	Implantable inorganic material	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	07/08/2023	07/08/2023

									A071102	Failure to Interrogate	07/08/2023	07/08/2023
							F1901	Additional Surgery	A070504	Premature Discharge of Battery	07/08/2023	07/08/2023
									A071102	Failure to Interrogate	07/08/2023	07/08/2023
					CT314	Electrically powered	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	07/08/2023	07/08/2023
									A071102	Failure to Interrogate	07/08/2023	07/08/2023
							F1901	Additional Surgery	A070504	Premature Discharge of Battery	07/08/2023	07/08/2023
									A071102	Failure to Interrogate	07/08/2023	07/08/2023
					CT321	Long-term surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	07/08/2023	07/08/2023
									A071102	Failure to Interrogate	07/08/2023	07/08/2023
							F1901	Additional Surgery	A070504	Premature Discharge of Battery	07/08/2023	07/08/2023
									A071102	Failure to Interrogate	07/08/2023	07/08/2023
					CT449	Generators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	07/08/2023	07/08/2023
									A071102	Failure to Interrogate	07/08/2023	07/08/2023
							F1901	Additional Surgery	A070504	Premature Discharge of Battery	07/08/2023	07/08/2023
									A071102	Failure to Interrogate	07/08/2023	07/08/2023
					CT981	Single use	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	07/08/2023	07/08/2023
									A071102	Failure to Interrogate	07/08/2023	07/08/2023
							F1901	Additional Surgery	A070504	Premature Discharge of Battery	07/08/2023	07/08/2023
									A071102	Failure to Interrogate	07/08/2023	07/08/2023
					CT983	Surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	07/08/2023	07/08/2023
									A071102	Failure to Interrogate	07/08/2023	07/08/2023
							F1901	Additional Surgery	A070504	Premature Discharge of Battery	07/08/2023	07/08/2023
									A071102	Failure to Interrogate	07/08/2023	07/08/2023
14101	2023		47265	Dual-chamber implantable pacemaker, rate-responsive	CT141	Electrodes/Leads and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	19/09/2023	19/09/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	19/09/2023	19/09/2023
					CT145	Cardiovascular devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	19/09/2023	19/09/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	19/09/2023	19/09/2023
					CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	19/09/2023	19/09/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	19/09/2023	19/09/2023
					CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	19/09/2023	19/09/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	19/09/2023	19/09/2023
					CT233	Surgical	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	19/09/2023	19/09/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	19/09/2023	19/09/2023
					CT2406	Implantable	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	19/09/2023	19/09/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	19/09/2023	19/09/2023
					CT2407	Active implantable	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	19/09/2023	19/09/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	19/09/2023	19/09/2023
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	19/09/2023	19/09/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	19/09/2023	19/09/2023
					CT2727	Active implant devices/ancillaries	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	19/09/2023	19/09/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	19/09/2023	19/09/2023
					CT2750	Implantable cardiac electrical stimulators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	19/09/2023	19/09/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	19/09/2023	19/09/2023
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	19/09/2023	19/09/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	19/09/2023	19/09/2023
					CT2985	Active	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	19/09/2023	19/09/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	19/09/2023	19/09/2023
					CT3001	Implantable metal	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	19/09/2023	19/09/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	19/09/2023	19/09/2023
					CT3002	Implantable synthetic polymer	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	19/09/2023	19/09/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	19/09/2023	19/09/2023
					CT3006	Implantable inorganic material	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	19/09/2023	19/09/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	19/09/2023	19/09/2023
					CT314	Electrically powered	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	19/09/2023	19/09/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	19/09/2023	19/09/2023
					CT321	Long-term surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	19/09/2023	19/09/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	19/09/2023	19/09/2023
					CT449	Generators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	19/09/2023	19/09/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	19/09/2023	19/09/2023
					CT981	Single use	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	19/09/2023	19/09/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	19/09/2023	19/09/2023
					CT983	Surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	19/09/2023	19/09/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	19/09/2023	19/09/2023
14102	2023	All other reportable incidents	47265	Dual-chamber implantable pacemaker, rate-responsive	CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A2501	Prophylactic Removal Due to Corrective Action	04/06/2023	04/06/2023
							F1901	Additional Surgery	A2501	Prophylactic Removal Due to Corrective Action	04/06/2023	04/06/2023
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A2501	Prophylactic Removal Due to Corrective Action	04/06/2023	04/06/2023
							F1901	Additional Surgery	A2501	Prophylactic Removal Due to Corrective Action	04/06/2023	04/06/2023
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A2501	Prophylactic Removal Due to Corrective Action	04/06/2023	04/06/2023
							F1901	Additional Surgery	A2501	Prophylactic Removal Due to Corrective Action	04/06/2023	04/06/2023
14103	2023	All other reportable incidents	47265	Dual-chamber implantable pacemaker, rate-responsive	CT141	Electrodes/Leads and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A071102	Failure to Interrogate		
									A071204	No Pacing		
							F1901	Additional Surgery	A071102	Failure to Interrogate		
									A071204	No Pacing		
					CT145	Cardiovascular devices	E2403	No Clinical Signs, Symptoms or Conditions	A071102	Failure to Interrogate		
									A071204	No Pacing		
							F1901	Additional Surgery	A071102	Failure to Interrogate		
									A071204	No Pacing		
					CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A071102	Failure to Interrogate		
									A071204	No Pacing		
							F1901	Additional Surgery	A071102	Failure to Interrogate		
									A071204	No Pacing		
					CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A071102	Failure to Interrogate		
									A071204	No Pacing		
							F1901	Additional Surgery	A071102	Failure to Interrogate		
									A071204	No Pacing		
					CT233	Surgical	E2403	No Clinical Signs, Symptoms or Conditions	A071102	Failure to Interrogate		
									A071204	No Pacing		
							F1901	Additional Surgery	A071102	Failure to Interrogate		

									A071204	No Pacing			
					CT2406	Implantable	E2403	No Clinical Signs, Symptoms or Conditions	A071102	Failure to Interrogate			
									A071204	No Pacing			
							F1901	Additional Surgery	A071102	Failure to Interrogate			
									A071204	No Pacing			
					CT2407	Active implantable	E2403	No Clinical Signs, Symptoms or Conditions	A071102	Failure to Interrogate			
									A071204	No Pacing			
							F1901	Additional Surgery	A071102	Failure to Interrogate			
									A071204	No Pacing			
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A071102	Failure to Interrogate			
									A071204	No Pacing			
							F1901	Additional Surgery	A071102	Failure to Interrogate			
									A071204	No Pacing			
					CT2727	Active implant devices/ancillaries	E2403	No Clinical Signs, Symptoms or Conditions	A071102	Failure to Interrogate			
									A071204	No Pacing			
							F1901	Additional Surgery	A071102	Failure to Interrogate			
									A071204	No Pacing			
					CT2750	Implantable cardiac electrical stimulators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A071102	Failure to Interrogate			
									A071204	No Pacing			
							F1901	Additional Surgery	A071102	Failure to Interrogate			
									A071204	No Pacing			
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A071102	Failure to Interrogate			
									A071204	No Pacing			
							F1901	Additional Surgery	A071102	Failure to Interrogate			
									A071204	No Pacing			
					CT2985	Active	E2403	No Clinical Signs, Symptoms or Conditions	A071102	Failure to Interrogate			
									A071204	No Pacing			
							F1901	Additional Surgery	A071102	Failure to Interrogate			
									A071204	No Pacing			
					CT3001	Implantable metal	E2403	No Clinical Signs, Symptoms or Conditions	A071102	Failure to Interrogate			
									A071204	No Pacing			
							F1901	Additional Surgery	A071102	Failure to Interrogate			
									A071204	No Pacing			
					CT3002	Implantable synthetic polymer	E2403	No Clinical Signs, Symptoms or Conditions	A071102	Failure to Interrogate			
									A071204	No Pacing			
							F1901	Additional Surgery	A071102	Failure to Interrogate			
									A071204	No Pacing			
					CT3006	Implantable inorganic material	E2403	No Clinical Signs, Symptoms or Conditions	A071102	Failure to Interrogate			
									A071204	No Pacing			
							F1901	Additional Surgery	A071102	Failure to Interrogate			
									A071204	No Pacing			
					CT314	Electrically powered	E2403	No Clinical Signs, Symptoms or Conditions	A071102	Failure to Interrogate			
									A071204	No Pacing			
							F1901	Additional Surgery	A071102	Failure to Interrogate			
									A071204	No Pacing			
					CT321	Long-term surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A071102	Failure to Interrogate			
									A071204	No Pacing			
							F1901	Additional Surgery	A071102	Failure to Interrogate			
									A071204	No Pacing			
					CT449	Generators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A071102	Failure to Interrogate			
									A071204	No Pacing			
							F1901	Additional Surgery	A071102	Failure to Interrogate			
									A071204	No Pacing			
					CT981	Single use	E2403	No Clinical Signs, Symptoms or Conditions	A071102	Failure to Interrogate			
									A071204	No Pacing			
							F1901	Additional Surgery	A071102	Failure to Interrogate			
									A071204	No Pacing			
					CT983	Surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A071102	Failure to Interrogate			
									A071204	No Pacing			
							F1901	Additional Surgery	A071102	Failure to Interrogate			
									A071204	No Pacing			
14104	2023		47265	Dual-chamber implantable pacemaker, rate-responsive	CT141	Electrodes/Leads and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	19/09/2023	19/09/2023	
							F1901	Additional Surgery	A24	No Apparent Device Problem	19/09/2023	19/09/2023	
					CT145	Cardiovascular devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	19/09/2023	19/09/2023	
							F1901	Additional Surgery	A24	No Apparent Device Problem	19/09/2023	19/09/2023	
					CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	19/09/2023	19/09/2023	
							F1901	Additional Surgery	A24	No Apparent Device Problem	19/09/2023	19/09/2023	
					CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	19/09/2023	19/09/2023	
							F1901	Additional Surgery	A24	No Apparent Device Problem	19/09/2023	19/09/2023	
					CT233	Surgical	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	19/09/2023	19/09/2023	
							F1901	Additional Surgery	A24	No Apparent Device Problem	19/09/2023	19/09/2023	
					CT2406	Implantable	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	19/09/2023	19/09/2023	
							F1901	Additional Surgery	A24	No Apparent Device Problem	19/09/2023	19/09/2023	
					CT2407	Active implantable	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	19/09/2023	19/09/2023	
							F1901	Additional Surgery	A24	No Apparent Device Problem	19/09/2023	19/09/2023	
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	19/09/2023	19/09/2023	
							F1901	Additional Surgery	A24	No Apparent Device Problem	19/09/2023	19/09/2023	
					CT2727	Active implant devices/ancillaries	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	19/09/2023	19/09/2023	
							F1901	Additional Surgery	A24	No Apparent Device Problem	19/09/2023	19/09/2023	
					CT2750	Implantable cardiac electrical stimulators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	19/09/2023	19/09/2023	
							F1901	Additional Surgery	A24	No Apparent Device Problem	19/09/2023	19/09/2023	
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	19/09/2023	19/09/2023	
							F1901	Additional Surgery	A24	No Apparent Device Problem	19/09/2023	19/09/2023	
					CT2985	Active	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	19/09/2023	19/09/2023	
							F1901	Additional Surgery	A24	No Apparent Device Problem	19/09/2023	19/09/2023	
					CT3001	Implantable metal	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	19/09/2023	19/09/2023	
							F1901	Additional Surgery	A24	No Apparent Device Problem	19/09/2023	19/09/2023	
					CT3002	Implantable synthetic polymer	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	19/09/2023	19/09/2023	

						F1901	Additional Surgery	A24	No Apparent Device Problem	19/09/2023	19/09/2023
				CT3006	Implantable inorganic material	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	19/09/2023	19/09/2023
						F1901	Additional Surgery	A24	No Apparent Device Problem	19/09/2023	19/09/2023
				CT314	Electrically powered	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	19/09/2023	19/09/2023
						F1901	Additional Surgery	A24	No Apparent Device Problem	19/09/2023	19/09/2023
				CT321	Long-term surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	19/09/2023	19/09/2023
						F1901	Additional Surgery	A24	No Apparent Device Problem	19/09/2023	19/09/2023
				CT449	Generators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	19/09/2023	19/09/2023
						F1901	Additional Surgery	A24	No Apparent Device Problem	19/09/2023	19/09/2023
				CT981	Single use	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	19/09/2023	19/09/2023
						F1901	Additional Surgery	A24	No Apparent Device Problem	19/09/2023	19/09/2023
				CT983	Surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	19/09/2023	19/09/2023
						F1901	Additional Surgery	A24	No Apparent Device Problem	19/09/2023	19/09/2023
14105	2023		47265	Dual-chamber implantable pacemaker, rate-responsive	CT141	Electrodes/Leads and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	20/09/2023	20/09/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	20/09/2023	20/09/2023
					CT145	Cardiovascular devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	20/09/2023	20/09/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	20/09/2023	20/09/2023
					CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	20/09/2023	20/09/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	20/09/2023	20/09/2023
					CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	20/09/2023	20/09/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	20/09/2023	20/09/2023
					CT233	Surgical	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	20/09/2023	20/09/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	20/09/2023	20/09/2023
					CT2406	Implantable	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	20/09/2023	20/09/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	20/09/2023	20/09/2023
					CT2407	Active implantable	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	20/09/2023	20/09/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	20/09/2023	20/09/2023
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	20/09/2023	20/09/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	20/09/2023	20/09/2023
					CT2727	Active implant devices/ancillaries	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	20/09/2023	20/09/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	20/09/2023	20/09/2023
					CT2750	Implantable cardiac electrical stimulators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	20/09/2023	20/09/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	20/09/2023	20/09/2023
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	20/09/2023	20/09/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	20/09/2023	20/09/2023
					CT2985	Active	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	20/09/2023	20/09/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	20/09/2023	20/09/2023
					CT3001	Implantable metal	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	20/09/2023	20/09/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	20/09/2023	20/09/2023
					CT3002	Implantable synthetic polymer	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	20/09/2023	20/09/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	20/09/2023	20/09/2023
					CT3006	Implantable inorganic material	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	20/09/2023	20/09/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	20/09/2023	20/09/2023
					CT314	Electrically powered	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	20/09/2023	20/09/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	20/09/2023	20/09/2023
					CT321	Long-term surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	20/09/2023	20/09/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	20/09/2023	20/09/2023
					CT449	Generators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	20/09/2023	20/09/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	20/09/2023	20/09/2023
					CT981	Single use	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	20/09/2023	20/09/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	20/09/2023	20/09/2023
					CT983	Surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	20/09/2023	20/09/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	20/09/2023	20/09/2023
14106	2023	All other reportable incidents	17882	Physiologic-monitoring defibrillation system	CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A0401	Break		
							F27	Problem Identified During Non-Clinical Procedure	A0401	Break		
14107	2023	All other reportable incidents	47265	Dual-chamber implantable pacemaker, rate-responsive	CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem		
							F26	No Health Consequences or Impact	A0705	Battery Problem		
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem		
							F26	No Health Consequences or Impact	A0705	Battery Problem		
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem		
							F26	No Health Consequences or Impact	A0705	Battery Problem		
14108	2023	All other reportable incidents	47265	Dual-chamber implantable pacemaker, rate-responsive	CT141	Electrodes/Leads and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070101	Failure to Capture		
									A070102	High Capture Threshold		
									A072201	High Impedance		
									A090801	Signal Artifact/Noise		
							F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A070101	Failure to Capture		
									A070102	High Capture Threshold		
									A072201	High Impedance		
									A090801	Signal Artifact/Noise		
							F2203	Imaging Required	A070101	Failure to Capture		
									A070102	High Capture Threshold		
									A072201	High Impedance		
									A090801	Signal Artifact/Noise		
					CT145	Cardiovascular devices	E2403	No Clinical Signs, Symptoms or Conditions	A070101	Failure to Capture		
									A070102	High Capture Threshold		
									A072201	High Impedance		
									A090801	Signal Artifact/Noise		
							F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A070101	Failure to Capture		
									A070102	High Capture Threshold		
									A072201	High Impedance		
									A090801	Signal Artifact/Noise		
							F2203	Imaging Required	A070101	Failure to Capture		
									A070102	High Capture Threshold		
									A072201	High Impedance		
									A090801	Signal Artifact/Noise		
					CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070101	Failure to Capture		
									A070102	High Capture Threshold		
									A072201	High Impedance		

									A090801	Signal Artifact/Noise		
							F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A070101	Failure to Capture		
									A070102	High Capture Threshold		
									A072201	High Impedance		
									A090801	Signal Artifact/Noise		
							F2203	Imaging Required	A070101	Failure to Capture		
									A070102	High Capture Threshold		
									A072201	High Impedance		
									A090801	Signal Artifact/Noise		
					CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A070101	Failure to Capture		
									A070102	High Capture Threshold		
									A072201	High Impedance		
									A090801	Signal Artifact/Noise		
							F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A070101	Failure to Capture		
									A070102	High Capture Threshold		
									A072201	High Impedance		
									A090801	Signal Artifact/Noise		
							F2203	Imaging Required	A070101	Failure to Capture		
									A070102	High Capture Threshold		
									A072201	High Impedance		
									A090801	Signal Artifact/Noise		
					CT233	Surgical	E2403	No Clinical Signs, Symptoms or Conditions	A070101	Failure to Capture		
									A070102	High Capture Threshold		
									A072201	High Impedance		
									A090801	Signal Artifact/Noise		
							F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A070101	Failure to Capture		
									A070102	High Capture Threshold		
									A072201	High Impedance		
									A090801	Signal Artifact/Noise		
							F2203	Imaging Required	A070101	Failure to Capture		
									A070102	High Capture Threshold		
									A072201	High Impedance		
									A090801	Signal Artifact/Noise		
					CT2406	Implantable	E2403	No Clinical Signs, Symptoms or Conditions	A070101	Failure to Capture		
									A070102	High Capture Threshold		
									A072201	High Impedance		
									A090801	Signal Artifact/Noise		
							F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A070101	Failure to Capture		
									A070102	High Capture Threshold		
									A072201	High Impedance		
									A090801	Signal Artifact/Noise		
							F2203	Imaging Required	A070101	Failure to Capture		
									A070102	High Capture Threshold		
									A072201	High Impedance		
									A090801	Signal Artifact/Noise		
					CT2407	Active implantable	E2403	No Clinical Signs, Symptoms or Conditions	A070101	Failure to Capture		
									A070102	High Capture Threshold		
									A072201	High Impedance		
									A090801	Signal Artifact/Noise		
							F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A070101	Failure to Capture		
									A070102	High Capture Threshold		
									A072201	High Impedance		
									A090801	Signal Artifact/Noise		
							F2203	Imaging Required	A070101	Failure to Capture		
									A070102	High Capture Threshold		
									A072201	High Impedance		
									A090801	Signal Artifact/Noise		
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A070101	Failure to Capture		
									A070102	High Capture Threshold		
									A072201	High Impedance		
									A090801	Signal Artifact/Noise		
							F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A070101	Failure to Capture		
									A070102	High Capture Threshold		
									A072201	High Impedance		
									A090801	Signal Artifact/Noise		
							F2203	Imaging Required	A070101	Failure to Capture		
									A070102	High Capture Threshold		
									A072201	High Impedance		
									A090801	Signal Artifact/Noise		
					CT2727	Active implant devices/ancillaries	E2403	No Clinical Signs, Symptoms or Conditions	A070101	Failure to Capture		
									A070102	High Capture Threshold		
									A072201	High Impedance		
									A090801	Signal Artifact/Noise		
							F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A070101	Failure to Capture		
									A070102	High Capture Threshold		
									A072201	High Impedance		
									A090801	Signal Artifact/Noise		
							F2203	Imaging Required	A070101	Failure to Capture		
									A070102	High Capture Threshold		
									A072201	High Impedance		
									A090801	Signal Artifact/Noise		
					CT2750	Implantable cardiac electrical stimulators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070101	Failure to Capture		
									A070102	High Capture Threshold		
									A072201	High Impedance		
									A090801	Signal Artifact/Noise		
							F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A070101	Failure to Capture		
									A070102	High Capture Threshold		
									A072201	High Impedance		

								A090801	Signal Artifact/Noise			
						F2203	Imaging Required	A070101	Failure to Capture			
								A070102	High Capture Threshold			
								A072201	High Impedance			
								A090801	Signal Artifact/Noise			
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A070101	Failure to Capture		
									A070102	High Capture Threshold		
									A072201	High Impedance		
									A090801	Signal Artifact/Noise		
							F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A070101	Failure to Capture		
									A070102	High Capture Threshold		
									A072201	High Impedance		
									A090801	Signal Artifact/Noise		
							F2203	Imaging Required	A070101	Failure to Capture		
									A070102	High Capture Threshold		
									A072201	High Impedance		
									A090801	Signal Artifact/Noise		
					CT2985	Active	E2403	No Clinical Signs, Symptoms or Conditions	A070101	Failure to Capture		
									A070102	High Capture Threshold		
									A072201	High Impedance		
									A090801	Signal Artifact/Noise		
							F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A070101	Failure to Capture		
									A070102	High Capture Threshold		
									A072201	High Impedance		
									A090801	Signal Artifact/Noise		
							F2203	Imaging Required	A070101	Failure to Capture		
									A070102	High Capture Threshold		
									A072201	High Impedance		
									A090801	Signal Artifact/Noise		
					CT3001	Implantable metal	E2403	No Clinical Signs, Symptoms or Conditions	A070101	Failure to Capture		
									A070102	High Capture Threshold		
									A072201	High Impedance		
									A090801	Signal Artifact/Noise		
							F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A070101	Failure to Capture		
									A070102	High Capture Threshold		
									A072201	High Impedance		
									A090801	Signal Artifact/Noise		
							F2203	Imaging Required	A070101	Failure to Capture		
									A070102	High Capture Threshold		
									A072201	High Impedance		
									A090801	Signal Artifact/Noise		
					CT3002	Implantable synthetic polymer	E2403	No Clinical Signs, Symptoms or Conditions	A070101	Failure to Capture		
									A070102	High Capture Threshold		
									A072201	High Impedance		
									A090801	Signal Artifact/Noise		
							F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A070101	Failure to Capture		
									A070102	High Capture Threshold		
									A072201	High Impedance		
									A090801	Signal Artifact/Noise		
							F2203	Imaging Required	A070101	Failure to Capture		
									A070102	High Capture Threshold		
									A072201	High Impedance		
									A090801	Signal Artifact/Noise		
					CT3006	Implantable inorganic material	E2403	No Clinical Signs, Symptoms or Conditions	A070101	Failure to Capture		
									A070102	High Capture Threshold		
									A072201	High Impedance		
									A090801	Signal Artifact/Noise		
							F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A070101	Failure to Capture		
									A070102	High Capture Threshold		
									A072201	High Impedance		
									A090801	Signal Artifact/Noise		
							F2203	Imaging Required	A070101	Failure to Capture		
									A070102	High Capture Threshold		
									A072201	High Impedance		
									A090801	Signal Artifact/Noise		
					CT314	Electrically powered	E2403	No Clinical Signs, Symptoms or Conditions	A070101	Failure to Capture		
									A070102	High Capture Threshold		
									A072201	High Impedance		
									A090801	Signal Artifact/Noise		
							F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A070101	Failure to Capture		
									A070102	High Capture Threshold		
									A072201	High Impedance		
									A090801	Signal Artifact/Noise		
							F2203	Imaging Required	A070101	Failure to Capture		
									A070102	High Capture Threshold		
									A072201	High Impedance		
									A090801	Signal Artifact/Noise		
					CT321	Long-term surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A070101	Failure to Capture		
									A070102	High Capture Threshold		
									A072201	High Impedance		
									A090801	Signal Artifact/Noise		
							F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A070101	Failure to Capture		
									A070102	High Capture Threshold		
									A072201	High Impedance		
									A090801	Signal Artifact/Noise		
							F2203	Imaging Required	A070101	Failure to Capture		
									A070102	High Capture Threshold		
									A072201	High Impedance		

									A090801	Signal Artifact/Noise		
					CT449	Generators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070101	Failure to Capture		
									A070102	High Capture Threshold		
									A072201	High Impedance		
									A090801	Signal Artifact/Noise		
							F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A070101	Failure to Capture		
									A070102	High Capture Threshold		
									A072201	High Impedance		
									A090801	Signal Artifact/Noise		
							F2203	Imaging Required	A070101	Failure to Capture		
									A070102	High Capture Threshold		
									A072201	High Impedance		
									A090801	Signal Artifact/Noise		
					CT981	Single use	E2403	No Clinical Signs, Symptoms or Conditions	A070101	Failure to Capture		
									A070102	High Capture Threshold		
									A072201	High Impedance		
									A090801	Signal Artifact/Noise		
							F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A070101	Failure to Capture		
									A070102	High Capture Threshold		
									A072201	High Impedance		
									A090801	Signal Artifact/Noise		
							F2203	Imaging Required	A070101	Failure to Capture		
									A070102	High Capture Threshold		
									A072201	High Impedance		
									A090801	Signal Artifact/Noise		
					CT983	Surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A070101	Failure to Capture		
									A070102	High Capture Threshold		
									A072201	High Impedance		
									A090801	Signal Artifact/Noise		
							F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A070101	Failure to Capture		
									A070102	High Capture Threshold		
									A072201	High Impedance		
									A090801	Signal Artifact/Noise		
							F2203	Imaging Required	A070101	Failure to Capture		
									A070102	High Capture Threshold		
									A072201	High Impedance		
									A090801	Signal Artifact/Noise		
14109	2023	All other reportable incidents	47265	Dual-chamber implantable pacemaker, rate-responsive	CT141	Electrodes/Leads and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A0908	Incorrect, Inadequate or Imprecise Result or Readings	25/05/2023	25/05/2023
							F08	Hospitalization or Prolonged Hospitalization	A0908	Incorrect, Inadequate or Imprecise Result or Readings	25/05/2023	25/05/2023
							F1905	Device Revision or Replacement	A0908	Incorrect, Inadequate or Imprecise Result or Readings	25/05/2023	25/05/2023
					CT145	Cardiovascular devices	E2403	No Clinical Signs, Symptoms or Conditions	A0908	Incorrect, Inadequate or Imprecise Result or Readings	25/05/2023	25/05/2023
							F08	Hospitalization or Prolonged Hospitalization	A0908	Incorrect, Inadequate or Imprecise Result or Readings	25/05/2023	25/05/2023
							F1905	Device Revision or Replacement	A0908	Incorrect, Inadequate or Imprecise Result or Readings	25/05/2023	25/05/2023
					CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A0908	Incorrect, Inadequate or Imprecise Result or Readings	25/05/2023	25/05/2023
							F08	Hospitalization or Prolonged Hospitalization	A0908	Incorrect, Inadequate or Imprecise Result or Readings	25/05/2023	25/05/2023
							F1905	Device Revision or Replacement	A0908	Incorrect, Inadequate or Imprecise Result or Readings	25/05/2023	25/05/2023
					CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A0908	Incorrect, Inadequate or Imprecise Result or Readings	25/05/2023	25/05/2023
							F08	Hospitalization or Prolonged Hospitalization	A0908	Incorrect, Inadequate or Imprecise Result or Readings	25/05/2023	25/05/2023
							F1905	Device Revision or Replacement	A0908	Incorrect, Inadequate or Imprecise Result or Readings	25/05/2023	25/05/2023
					CT233	Surgical	E2403	No Clinical Signs, Symptoms or Conditions	A0908	Incorrect, Inadequate or Imprecise Result or Readings	25/05/2023	25/05/2023
							F08	Hospitalization or Prolonged Hospitalization	A0908	Incorrect, Inadequate or Imprecise Result or Readings	25/05/2023	25/05/2023
							F1905	Device Revision or Replacement	A0908	Incorrect, Inadequate or Imprecise Result or Readings	25/05/2023	25/05/2023
					CT2406	Implantable	E2403	No Clinical Signs, Symptoms or Conditions	A0908	Incorrect, Inadequate or Imprecise Result or Readings	25/05/2023	25/05/2023
							F08	Hospitalization or Prolonged Hospitalization	A0908	Incorrect, Inadequate or Imprecise Result or Readings	25/05/2023	25/05/2023
							F1905	Device Revision or Replacement	A0908	Incorrect, Inadequate or Imprecise Result or Readings	25/05/2023	25/05/2023
					CT2407	Active implantable	E2403	No Clinical Signs, Symptoms or Conditions	A0908	Incorrect, Inadequate or Imprecise Result or Readings	25/05/2023	25/05/2023
							F08	Hospitalization or Prolonged Hospitalization	A0908	Incorrect, Inadequate or Imprecise Result or Readings	25/05/2023	25/05/2023
							F1905	Device Revision or Replacement	A0908	Incorrect, Inadequate or Imprecise Result or Readings	25/05/2023	25/05/2023
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A0908	Incorrect, Inadequate or Imprecise Result or Readings	25/05/2023	25/05/2023
							F08	Hospitalization or Prolonged Hospitalization	A0908	Incorrect, Inadequate or Imprecise Result or Readings	25/05/2023	25/05/2023
							F1905	Device Revision or Replacement	A0908	Incorrect, Inadequate or Imprecise Result or Readings	25/05/2023	25/05/2023
					CT2727	Active implant devices/ancillaries	E2403	No Clinical Signs, Symptoms or Conditions	A0908	Incorrect, Inadequate or Imprecise Result or Readings	25/05/2023	25/05/2023
							F08	Hospitalization or Prolonged Hospitalization	A0908	Incorrect, Inadequate or Imprecise Result or Readings	25/05/2023	25/05/2023
							F1905	Device Revision or Replacement	A0908	Incorrect, Inadequate or Imprecise Result or Readings	25/05/2023	25/05/2023
					CT2750	Implantable cardiac electrical stimulators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A0908	Incorrect, Inadequate or Imprecise Result or Readings	25/05/2023	25/05/2023
							F08	Hospitalization or Prolonged Hospitalization	A0908	Incorrect, Inadequate or Imprecise Result or Readings	25/05/2023	25/05/2023

							F1905	Device Revision or Replacement	A0908	Incorrect, Inadequate or Imprecise Result or Readings	25/05/2023	25/05/2023
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A0908	Incorrect, Inadequate or Imprecise Result or Readings	25/05/2023	25/05/2023
							F08	Hospitalization or Prolonged Hospitalization	A0908	Incorrect, Inadequate or Imprecise Result or Readings	25/05/2023	25/05/2023
							F1905	Device Revision or Replacement	A0908	Incorrect, Inadequate or Imprecise Result or Readings	25/05/2023	25/05/2023
					CT2985	Active	E2403	No Clinical Signs, Symptoms or Conditions	A0908	Incorrect, Inadequate or Imprecise Result or Readings	25/05/2023	25/05/2023
							F08	Hospitalization or Prolonged Hospitalization	A0908	Incorrect, Inadequate or Imprecise Result or Readings	25/05/2023	25/05/2023
							F1905	Device Revision or Replacement	A0908	Incorrect, Inadequate or Imprecise Result or Readings	25/05/2023	25/05/2023
					CT3001	Implantable metal	E2403	No Clinical Signs, Symptoms or Conditions	A0908	Incorrect, Inadequate or Imprecise Result or Readings	25/05/2023	25/05/2023
							F08	Hospitalization or Prolonged Hospitalization	A0908	Incorrect, Inadequate or Imprecise Result or Readings	25/05/2023	25/05/2023
							F1905	Device Revision or Replacement	A0908	Incorrect, Inadequate or Imprecise Result or Readings	25/05/2023	25/05/2023
					CT3002	Implantable synthetic polymer	E2403	No Clinical Signs, Symptoms or Conditions	A0908	Incorrect, Inadequate or Imprecise Result or Readings	25/05/2023	25/05/2023
							F08	Hospitalization or Prolonged Hospitalization	A0908	Incorrect, Inadequate or Imprecise Result or Readings	25/05/2023	25/05/2023
							F1905	Device Revision or Replacement	A0908	Incorrect, Inadequate or Imprecise Result or Readings	25/05/2023	25/05/2023
					CT3006	Implantable inorganic material	E2403	No Clinical Signs, Symptoms or Conditions	A0908	Incorrect, Inadequate or Imprecise Result or Readings	25/05/2023	25/05/2023
							F08	Hospitalization or Prolonged Hospitalization	A0908	Incorrect, Inadequate or Imprecise Result or Readings	25/05/2023	25/05/2023
							F1905	Device Revision or Replacement	A0908	Incorrect, Inadequate or Imprecise Result or Readings	25/05/2023	25/05/2023
					CT314	Electrically powered	E2403	No Clinical Signs, Symptoms or Conditions	A0908	Incorrect, Inadequate or Imprecise Result or Readings	25/05/2023	25/05/2023
							F08	Hospitalization or Prolonged Hospitalization	A0908	Incorrect, Inadequate or Imprecise Result or Readings	25/05/2023	25/05/2023
							F1905	Device Revision or Replacement	A0908	Incorrect, Inadequate or Imprecise Result or Readings	25/05/2023	25/05/2023
					CT321	Long-term surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A0908	Incorrect, Inadequate or Imprecise Result or Readings	25/05/2023	25/05/2023
							F08	Hospitalization or Prolonged Hospitalization	A0908	Incorrect, Inadequate or Imprecise Result or Readings	25/05/2023	25/05/2023
							F1905	Device Revision or Replacement	A0908	Incorrect, Inadequate or Imprecise Result or Readings	25/05/2023	25/05/2023
					CT449	Generators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A0908	Incorrect, Inadequate or Imprecise Result or Readings	25/05/2023	25/05/2023
							F08	Hospitalization or Prolonged Hospitalization	A0908	Incorrect, Inadequate or Imprecise Result or Readings	25/05/2023	25/05/2023
							F1905	Device Revision or Replacement	A0908	Incorrect, Inadequate or Imprecise Result or Readings	25/05/2023	25/05/2023
					CT981	Single use	E2403	No Clinical Signs, Symptoms or Conditions	A0908	Incorrect, Inadequate or Imprecise Result or Readings	25/05/2023	25/05/2023
							F08	Hospitalization or Prolonged Hospitalization	A0908	Incorrect, Inadequate or Imprecise Result or Readings	25/05/2023	25/05/2023
							F1905	Device Revision or Replacement	A0908	Incorrect, Inadequate or Imprecise Result or Readings	25/05/2023	25/05/2023
					CT983	Surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A0908	Incorrect, Inadequate or Imprecise Result or Readings	25/05/2023	25/05/2023
							F08	Hospitalization or Prolonged Hospitalization	A0908	Incorrect, Inadequate or Imprecise Result or Readings	25/05/2023	25/05/2023
							F1905	Device Revision or Replacement	A0908	Incorrect, Inadequate or Imprecise Result or Readings	25/05/2023	25/05/2023
14110	2023	Death	60789	Intracardiac pacemaker	CT130	Catheters and associated devices	E2401	Insufficient Information	A0701	Capturing Problem		
									A1502	Positioning Problem		
							E2403	No Clinical Signs, Symptoms or Conditions	A0701	Capturing Problem		
									A1502	Positioning Problem		
							F02	Death	A0701	Capturing Problem		
									A1502	Positioning Problem		
							F26	No Health Consequences or Impact	A0701	Capturing Problem		
									A1502	Positioning Problem		
					CT145	Cardiovascular devices	E2401	Insufficient Information	A0701	Capturing Problem		
									A1502	Positioning Problem		
							E2403	No Clinical Signs, Symptoms or Conditions	A0701	Capturing Problem		
									A1502	Positioning Problem		
							F02	Death	A0701	Capturing Problem		
									A1502	Positioning Problem		
							F26	No Health Consequences or Impact	A0701	Capturing Problem		
									A1502	Positioning Problem		
					CT1684	Cardiac pacemakers and associated devices	E2401	Insufficient Information	A0701	Capturing Problem		
									A1502	Positioning Problem		
							E2403	No Clinical Signs, Symptoms or Conditions	A0701	Capturing Problem		
									A1502	Positioning Problem		
							F02	Death	A0701	Capturing Problem		
									A1502	Positioning Problem		
							F26	No Health Consequences or Impact	A0701	Capturing Problem		
									A1502	Positioning Problem		
					CT1685	Cardiac pacemakers	E2401	Insufficient Information	A0701	Capturing Problem		
									A1502	Positioning Problem		
							E2403	No Clinical Signs, Symptoms or Conditions	A0701	Capturing Problem		
									A1502	Positioning Problem		
							F02	Death	A0701	Capturing Problem		
									A1502	Positioning Problem		
							F26	No Health Consequences or Impact	A0701	Capturing Problem		
									A1502	Positioning Problem		
					CT233	Surgical	E2401	Insufficient Information	A0701	Capturing Problem		

								A1502	Positioning Problem		
						E2403	No Clinical Signs, Symptoms or Conditions	A0701	Capturing Problem		
								A1502	Positioning Problem		
						F02	Death	A0701	Capturing Problem		
								A1502	Positioning Problem		
						F26	No Health Consequences or Impact	A0701	Capturing Problem		
								A1502	Positioning Problem		
					CT2406	Implantable	E2401	Insufficient Information	A0701	Capturing Problem		
									A1502	Positioning Problem		
							E2403	No Clinical Signs, Symptoms or Conditions	A0701	Capturing Problem		
									A1502	Positioning Problem		
							F02	Death	A0701	Capturing Problem		
									A1502	Positioning Problem		
							F26	No Health Consequences or Impact	A0701	Capturing Problem		
									A1502	Positioning Problem		
					CT2407	Active implantable	E2401	Insufficient Information	A0701	Capturing Problem		
									A1502	Positioning Problem		
							E2403	No Clinical Signs, Symptoms or Conditions	A0701	Capturing Problem		
									A1502	Positioning Problem		
							F02	Death	A0701	Capturing Problem		
									A1502	Positioning Problem		
							F26	No Health Consequences or Impact	A0701	Capturing Problem		
									A1502	Positioning Problem		
					CT257	Implantable cardiac pacemakers	E2401	Insufficient Information	A0701	Capturing Problem		
									A1502	Positioning Problem		
							E2403	No Clinical Signs, Symptoms or Conditions	A0701	Capturing Problem		
									A1502	Positioning Problem		
							F02	Death	A0701	Capturing Problem		
									A1502	Positioning Problem		
							F26	No Health Consequences or Impact	A0701	Capturing Problem		
									A1502	Positioning Problem		
					CT2727	Active implant devices/ancillaries	E2401	Insufficient Information	A0701	Capturing Problem		
									A1502	Positioning Problem		
							E2403	No Clinical Signs, Symptoms or Conditions	A0701	Capturing Problem		
									A1502	Positioning Problem		
							F02	Death	A0701	Capturing Problem		
									A1502	Positioning Problem		
							F26	No Health Consequences or Impact	A0701	Capturing Problem		
									A1502	Positioning Problem		
					CT2750	Implantable cardiac electrical stimulators and associated devices	E2401	Insufficient Information	A0701	Capturing Problem		
									A1502	Positioning Problem		
							E2403	No Clinical Signs, Symptoms or Conditions	A0701	Capturing Problem		
									A1502	Positioning Problem		
							F02	Death	A0701	Capturing Problem		
									A1502	Positioning Problem		
							F26	No Health Consequences or Impact	A0701	Capturing Problem		
									A1502	Positioning Problem		
					CT2818	Catheterization/Cannulation/Fluid path devices	E2401	Insufficient Information	A0701	Capturing Problem		
									A1502	Positioning Problem		
							E2403	No Clinical Signs, Symptoms or Conditions	A0701	Capturing Problem		
									A1502	Positioning Problem		
							F02	Death	A0701	Capturing Problem		
									A1502	Positioning Problem		
							F26	No Health Consequences or Impact	A0701	Capturing Problem		
									A1502	Positioning Problem		
					CT2866	Implantable cardiac electrical stimulators	E2401	Insufficient Information	A0701	Capturing Problem		
									A1502	Positioning Problem		
							E2403	No Clinical Signs, Symptoms or Conditions	A0701	Capturing Problem		
									A1502	Positioning Problem		
							F02	Death	A0701	Capturing Problem		
									A1502	Positioning Problem		
							F26	No Health Consequences or Impact	A0701	Capturing Problem		
									A1502	Positioning Problem		
					CT2985	Active	E2401	Insufficient Information	A0701	Capturing Problem		
									A1502	Positioning Problem		
							E2403	No Clinical Signs, Symptoms or Conditions	A0701	Capturing Problem		
									A1502	Positioning Problem		
							F02	Death	A0701	Capturing Problem		
									A1502	Positioning Problem		
							F26	No Health Consequences or Impact	A0701	Capturing Problem		
									A1502	Positioning Problem		
					CT3001	Implantable metal	E2401	Insufficient Information	A0701	Capturing Problem		
									A1502	Positioning Problem		
							E2403	No Clinical Signs, Symptoms or Conditions	A0701	Capturing Problem		
									A1502	Positioning Problem		
							F02	Death	A0701	Capturing Problem		
									A1502	Positioning Problem		
							F26	No Health Consequences or Impact	A0701	Capturing Problem		
									A1502	Positioning Problem		
					CT3002	Implantable synthetic polymer	E2401	Insufficient Information	A0701	Capturing Problem		
									A1502	Positioning Problem		
							E2403	No Clinical Signs, Symptoms or Conditions	A0701	Capturing Problem		
									A1502	Positioning Problem		
							F02	Death	A0701	Capturing Problem		
									A1502	Positioning Problem		
							F26	No Health Consequences or Impact	A0701	Capturing Problem		
									A1502	Positioning Problem		
					CT3006	Implantable inorganic material	E2401	Insufficient Information	A0701	Capturing Problem		

									A1502	Positioning Problem		
							E2403	No Clinical Signs, Symptoms or Conditions	A0701	Capturing Problem		
									A1502	Positioning Problem		
							F02	Death	A0701	Capturing Problem		
									A1502	Positioning Problem		
							F26	No Health Consequences or Impact	A0701	Capturing Problem		
									A1502	Positioning Problem		
					CT314	Electrically powered	E2401	Insufficient Information	A0701	Capturing Problem		
									A1502	Positioning Problem		
							E2403	No Clinical Signs, Symptoms or Conditions	A0701	Capturing Problem		
									A1502	Positioning Problem		
							F02	Death	A0701	Capturing Problem		
									A1502	Positioning Problem		
							F26	No Health Consequences or Impact	A0701	Capturing Problem		
									A1502	Positioning Problem		
					CT321	Long-term surgical invasive	E2401	Insufficient Information	A0701	Capturing Problem		
									A1502	Positioning Problem		
							E2403	No Clinical Signs, Symptoms or Conditions	A0701	Capturing Problem		
									A1502	Positioning Problem		
							F02	Death	A0701	Capturing Problem		
									A1502	Positioning Problem		
							F26	No Health Consequences or Impact	A0701	Capturing Problem		
									A1502	Positioning Problem		
					CT981	Single use	E2401	Insufficient Information	A0701	Capturing Problem		
									A1502	Positioning Problem		
							E2403	No Clinical Signs, Symptoms or Conditions	A0701	Capturing Problem		
									A1502	Positioning Problem		
							F02	Death	A0701	Capturing Problem		
									A1502	Positioning Problem		
							F26	No Health Consequences or Impact	A0701	Capturing Problem		
									A1502	Positioning Problem		
					CT983	Surgical invasive	E2401	Insufficient Information	A0701	Capturing Problem		
									A1502	Positioning Problem		
							E2403	No Clinical Signs, Symptoms or Conditions	A0701	Capturing Problem		
									A1502	Positioning Problem		
							F02	Death	A0701	Capturing Problem		
									A1502	Positioning Problem		
							F26	No Health Consequences or Impact	A0701	Capturing Problem		
									A1502	Positioning Problem		
14111	2023	All other reportable incidents	35223	Endocardial/interventricular septal pacing lead	CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070909	Oversensing		
							F1901	Additional Surgery	A070909	Oversensing		
14112	2023	All other reportable incidents	47267	Single-chamber implantable pacemaker, rate-responsive	CT141	Electrodes/Leads and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
					CT145	Cardiovascular devices	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
					CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
					CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
					CT233	Surgical	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
					CT2406	Implantable	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
					CT2407	Active implantable	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
					CT2727	Active implant devices/ancillaries	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
					CT2750	Implantable cardiac electrical stimulators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
					CT2985	Active	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A0705	Battery Problem		

									A070504	Premature Discharge of Battery		
					CT3001	Implantable metal	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
					CT3002	Implantable synthetic polymer	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
					CT3006	Implantable inorganic material	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
					CT314	Electrically powered	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
					CT321	Long-term surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
					CT449	Generators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
					CT981	Single use	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
					CT983	Surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
14113	2023	All other reportable incidents	47263	Cardiac resynchronization therapy implantable pacemaker	CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A0908	Incorrect, Inadequate or Imprecise Result or Readings	30/12/2022	30/12/2022
									A1601	Device Alarm System	30/12/2022	30/12/2022
							F08	Hospitalization or Prolonged Hospitalization	A0908	Incorrect, Inadequate or Imprecise Result or Readings	30/12/2022	30/12/2022
									A1601	Device Alarm System	30/12/2022	30/12/2022
							F1905	Device Revision or Replacement	A0908	Incorrect, Inadequate or Imprecise Result or Readings	30/12/2022	30/12/2022
									A1601	Device Alarm System	30/12/2022	30/12/2022
					CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A0908	Incorrect, Inadequate or Imprecise Result or Readings	30/12/2022	30/12/2022
									A1601	Device Alarm System	30/12/2022	30/12/2022
							F08	Hospitalization or Prolonged Hospitalization	A0908	Incorrect, Inadequate or Imprecise Result or Readings	30/12/2022	30/12/2022
									A1601	Device Alarm System	30/12/2022	30/12/2022
							F1905	Device Revision or Replacement	A0908	Incorrect, Inadequate or Imprecise Result or Readings	30/12/2022	30/12/2022
									A1601	Device Alarm System	30/12/2022	30/12/2022
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A0908	Incorrect, Inadequate or Imprecise Result or Readings	30/12/2022	30/12/2022
									A1601	Device Alarm System	30/12/2022	30/12/2022
							F08	Hospitalization or Prolonged Hospitalization	A0908	Incorrect, Inadequate or Imprecise Result or Readings	30/12/2022	30/12/2022
									A1601	Device Alarm System	30/12/2022	30/12/2022
							F1905	Device Revision or Replacement	A0908	Incorrect, Inadequate or Imprecise Result or Readings	30/12/2022	30/12/2022
									A1601	Device Alarm System	30/12/2022	30/12/2022
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A0908	Incorrect, Inadequate or Imprecise Result or Readings	30/12/2022	30/12/2022
									A1601	Device Alarm System	30/12/2022	30/12/2022
							F08	Hospitalization or Prolonged Hospitalization	A0908	Incorrect, Inadequate or Imprecise Result or Readings	30/12/2022	30/12/2022
									A1601	Device Alarm System	30/12/2022	30/12/2022
							F1905	Device Revision or Replacement	A0908	Incorrect, Inadequate or Imprecise Result or Readings	30/12/2022	30/12/2022
									A1601	Device Alarm System	30/12/2022	30/12/2022
14114	2023	All other reportable incidents			CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070909	Oversensing		
							F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A070909	Oversensing		
							F23	Unexpected Medical Intervention	A070909	Oversensing		
14115	2023	All other reportable incidents	47205	Cardiac pulse generator programmer	CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A0203	Defective Device		
									A150105	Key or Button Unresponsive/not Working		
							F1908	Prolonged Surgery	A0203	Defective Device		
									A150105	Key or Button Unresponsive/not Working		
14116	2023	All other reportable incidents	46717	Pacing/defibrillation lead stylet	CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070102	High Capture Threshold	04/08/2023	04/08/2023
									A0722	Impedance Problem	04/08/2023	04/08/2023
									A072202	Low Impedance	04/08/2023	04/08/2023
							F08	Hospitalization or Prolonged Hospitalization	A070102	High Capture Threshold	04/08/2023	04/08/2023
									A0722	Impedance Problem	04/08/2023	04/08/2023
									A072202	Low Impedance	04/08/2023	04/08/2023
							F1905	Device Revision or Replacement	A070102	High Capture Threshold	04/08/2023	04/08/2023
									A0722	Impedance Problem	04/08/2023	04/08/2023
									A072202	Low Impedance	04/08/2023	04/08/2023
14117	2023		47265	Dual-chamber implantable pacemaker, rate-responsive	CT141	Electrodes/Leads and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	20/09/2023	20/09/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	20/09/2023	20/09/2023
					CT145	Cardiovascular devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	20/09/2023	20/09/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	20/09/2023	20/09/2023
					CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	20/09/2023	20/09/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	20/09/2023	20/09/2023

					CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	20/09/2023	20/09/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	20/09/2023	20/09/2023
					CT233	Surgical	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	20/09/2023	20/09/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	20/09/2023	20/09/2023
					CT2406	Implantable	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	20/09/2023	20/09/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	20/09/2023	20/09/2023
					CT2407	Active implantable	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	20/09/2023	20/09/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	20/09/2023	20/09/2023
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	20/09/2023	20/09/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	20/09/2023	20/09/2023
					CT2727	Active implant devices/ancillaries	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	20/09/2023	20/09/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	20/09/2023	20/09/2023
					CT2750	Implantable cardiac electrical stimulators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	20/09/2023	20/09/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	20/09/2023	20/09/2023
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	20/09/2023	20/09/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	20/09/2023	20/09/2023
					CT2985	Active	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	20/09/2023	20/09/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	20/09/2023	20/09/2023
					CT3001	Implantable metal	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	20/09/2023	20/09/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	20/09/2023	20/09/2023
					CT3002	Implantable synthetic polymer	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	20/09/2023	20/09/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	20/09/2023	20/09/2023
					CT3006	Implantable inorganic material	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	20/09/2023	20/09/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	20/09/2023	20/09/2023
					CT314	Electrically powered	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	20/09/2023	20/09/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	20/09/2023	20/09/2023
					CT321	Long-term surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	20/09/2023	20/09/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	20/09/2023	20/09/2023
					CT449	Generators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	20/09/2023	20/09/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	20/09/2023	20/09/2023
					CT981	Single use	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	20/09/2023	20/09/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	20/09/2023	20/09/2023
					CT983	Surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	20/09/2023	20/09/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	20/09/2023	20/09/2023
14118	2023		47265	Dual-chamber implantable pacemaker, rate-responsive	CT141	Electrodes/Leads and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	20/09/2023	20/09/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	20/09/2023	20/09/2023
					CT145	Cardiovascular devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	20/09/2023	20/09/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	20/09/2023	20/09/2023
					CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	20/09/2023	20/09/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	20/09/2023	20/09/2023
					CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	20/09/2023	20/09/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	20/09/2023	20/09/2023
					CT233	Surgical	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	20/09/2023	20/09/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	20/09/2023	20/09/2023
					CT2406	Implantable	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	20/09/2023	20/09/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	20/09/2023	20/09/2023
					CT2407	Active implantable	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	20/09/2023	20/09/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	20/09/2023	20/09/2023
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	20/09/2023	20/09/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	20/09/2023	20/09/2023
					CT2727	Active implant devices/ancillaries	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	20/09/2023	20/09/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	20/09/2023	20/09/2023
					CT2750	Implantable cardiac electrical stimulators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	20/09/2023	20/09/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	20/09/2023	20/09/2023
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	20/09/2023	20/09/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	20/09/2023	20/09/2023
					CT2985	Active	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	20/09/2023	20/09/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	20/09/2023	20/09/2023
					CT3001	Implantable metal	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	20/09/2023	20/09/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	20/09/2023	20/09/2023
					CT3002	Implantable synthetic polymer	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	20/09/2023	20/09/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	20/09/2023	20/09/2023
					CT3006	Implantable inorganic material	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	20/09/2023	20/09/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	20/09/2023	20/09/2023
					CT314	Electrically powered	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	20/09/2023	20/09/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	20/09/2023	20/09/2023
					CT321	Long-term surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	20/09/2023	20/09/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	20/09/2023	20/09/2023
					CT449	Generators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	20/09/2023	20/09/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	20/09/2023	20/09/2023
					CT981	Single use	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	20/09/2023	20/09/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	20/09/2023	20/09/2023
					CT983	Surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	20/09/2023	20/09/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	20/09/2023	20/09/2023
14119	2023		47265	Dual-chamber implantable pacemaker, rate-responsive	CT141	Electrodes/Leads and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	20/09/2023	20/09/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	20/09/2023	20/09/2023
					CT145	Cardiovascular devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	20/09/2023	20/09/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	20/09/2023	20/09/2023
					CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	20/09/2023	20/09/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	20/09/2023	20/09/2023
					CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	20/09/2023	20/09/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	20/09/2023	20/09/2023
					CT233	Surgical	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	20/09/2023	20/09/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	20/09/2023	20/09/2023
					CT2406	Implantable	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	20/09/2023	20/09/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	20/09/2023	20/09/2023
					CT2407	Active implantable	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	20/09/2023	20/09/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	20/09/2023	20/09/2023

					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	20/09/2023	20/09/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	20/09/2023	20/09/2023
					CT2727	Active implant devices/ancillaries	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	20/09/2023	20/09/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	20/09/2023	20/09/2023
					CT2750	Implantable cardiac electrical stimulators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	20/09/2023	20/09/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	20/09/2023	20/09/2023
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	20/09/2023	20/09/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	20/09/2023	20/09/2023
					CT2985	Active	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	20/09/2023	20/09/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	20/09/2023	20/09/2023
					CT3001	Implantable metal	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	20/09/2023	20/09/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	20/09/2023	20/09/2023
					CT3002	Implantable synthetic polymer	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	20/09/2023	20/09/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	20/09/2023	20/09/2023
					CT3006	Implantable inorganic material	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	20/09/2023	20/09/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	20/09/2023	20/09/2023
					CT314	Electrically powered	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	20/09/2023	20/09/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	20/09/2023	20/09/2023
					CT321	Long-term surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	20/09/2023	20/09/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	20/09/2023	20/09/2023
					CT449	Generators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	20/09/2023	20/09/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	20/09/2023	20/09/2023
					CT981	Single use	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	20/09/2023	20/09/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	20/09/2023	20/09/2023
					CT983	Surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	20/09/2023	20/09/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	20/09/2023	20/09/2023
14120	2023		35223	Endocardial/interventricular septal pacing lead	CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A25	No Apparent Adverse Event		
							F26	No Health Consequences or Impact	A25	No Apparent Adverse Event		
14121	2023	All other reportable incidents	47270	Cardiac resynchronization therapy implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A070603	Failure to Charge	06/06/2023	06/06/2023
							F1901	Additional Surgery	A070603	Failure to Charge	06/06/2023	06/06/2023
14122	2023	All other reportable incidents	17882	Physiologic-monitoring defibrillation system	CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A0908	Incorrect, Inadequate or Imprecise Result or Readings		
							F26	No Health Consequences or Impact	A0908	Incorrect, Inadequate or Imprecise Result or Readings		
14123	2023	All other reportable incidents	47263	Cardiac resynchronization therapy implantable pacemaker	CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A090801	Signal Artifact/Noise		
							F26	No Health Consequences or Impact	A090801	Signal Artifact/Noise		
					CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A090801	Signal Artifact/Noise		
							F26	No Health Consequences or Impact	A090801	Signal Artifact/Noise		
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A090801	Signal Artifact/Noise		
							F26	No Health Consequences or Impact	A090801	Signal Artifact/Noise		
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A090801	Signal Artifact/Noise		
							F26	No Health Consequences or Impact	A090801	Signal Artifact/Noise		
14124	2023	All other reportable incidents	47265	Dual-chamber implantable pacemaker, rate-responsive	CT141	Electrodes/Leads and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
					CT145	Cardiovascular devices	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
					CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
					CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
					CT233	Surgical	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
					CT2406	Implantable	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
					CT2407	Active implantable	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
					CT2727	Active implant devices/ancillaries	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
					CT2750	Implantable cardiac electrical stimulators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
					CT2985	Active	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		

						F26	No Health Consequences or Impact	A0705	Battery Problem			
								A070504	Premature Discharge of Battery			
					CT3001	Implantable metal	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
					CT3002	Implantable synthetic polymer	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
					CT3006	Implantable inorganic material	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
					CT314	Electrically powered	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
					CT321	Long-term surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
					CT449	Generators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
					CT981	Single use	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
					CT983	Surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
14125	2023	All other reportable incidents	35852	Single-chamber implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	07/12/2023	07/12/2023
									A0908	Incorrect, Inadequate or Imprecise Result or Readings	07/12/2023	07/12/2023
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	07/12/2023	07/12/2023
									A0908	Incorrect, Inadequate or Imprecise Result or Readings	07/12/2023	07/12/2023
							F1903	Device Explantation	A070504	Premature Discharge of Battery	07/12/2023	07/12/2023
									A0908	Incorrect, Inadequate or Imprecise Result or Readings	07/12/2023	07/12/2023
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	07/12/2023	07/12/2023
									A0908	Incorrect, Inadequate or Imprecise Result or Readings	07/12/2023	07/12/2023
14126	2023	All other reportable incidents	36241	Pacemaker/defibrillator lead	CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A072201	High Impedance		
							F26	No Health Consequences or Impact	A072201	High Impedance		
14127	2023	All other reportable incidents	47265	Dual-chamber implantable pacemaker, rate-responsive	CT141	Electrodes/Leads and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem		
							F24	Insufficient Health Impact Information	A0705	Battery Problem		
					CT145	Cardiovascular devices	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem		
							F24	Insufficient Health Impact Information	A0705	Battery Problem		
					CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem		
							F24	Insufficient Health Impact Information	A0705	Battery Problem		
					CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem		
							F24	Insufficient Health Impact Information	A0705	Battery Problem		
					CT233	Surgical	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem		
							F24	Insufficient Health Impact Information	A0705	Battery Problem		
					CT2406	Implantable	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem		
							F24	Insufficient Health Impact Information	A0705	Battery Problem		
					CT2407	Active implantable	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem		
							F24	Insufficient Health Impact Information	A0705	Battery Problem		
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem		
							F24	Insufficient Health Impact Information	A0705	Battery Problem		
					CT2727	Active implant devices/ancillaries	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem		
							F24	Insufficient Health Impact Information	A0705	Battery Problem		
					CT2750	Implantable cardiac electrical stimulators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem		
							F24	Insufficient Health Impact Information	A0705	Battery Problem		
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem		
							F24	Insufficient Health Impact Information	A0705	Battery Problem		
					CT2985	Active	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem		
							F24	Insufficient Health Impact Information	A0705	Battery Problem		
					CT3001	Implantable metal	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem		
							F24	Insufficient Health Impact Information	A0705	Battery Problem		
					CT3002	Implantable synthetic polymer	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem		
							F24	Insufficient Health Impact Information	A0705	Battery Problem		
					CT3006	Implantable inorganic material	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem		
							F24	Insufficient Health Impact Information	A0705	Battery Problem		
					CT314	Electrically powered	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem		
							F24	Insufficient Health Impact Information	A0705	Battery Problem		
					CT321	Long-term surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem		
							F24	Insufficient Health Impact Information	A0705	Battery Problem		
					CT449	Generators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem		
							F24	Insufficient Health Impact Information	A0705	Battery Problem		
					CT981	Single use	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem		
							F24	Insufficient Health Impact Information	A0705	Battery Problem		
					CT983	Surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem		
							F24	Insufficient Health Impact Information	A0705	Battery Problem		
14128	2023	All other reportable incidents	47265	Dual-chamber implantable pacemaker, rate-	CT141	Electrodes/Leads and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	30/08/2023	30/08/2023

				responsive			F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	30/08/2023	30/08/2023
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	30/08/2023	30/08/2023
					CT145	Cardiovascular devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	30/08/2023	30/08/2023
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	30/08/2023	30/08/2023
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	30/08/2023	30/08/2023
					CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	30/08/2023	30/08/2023
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	30/08/2023	30/08/2023
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	30/08/2023	30/08/2023
					CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	30/08/2023	30/08/2023
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	30/08/2023	30/08/2023
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	30/08/2023	30/08/2023
					CT233	Surgical	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	30/08/2023	30/08/2023
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	30/08/2023	30/08/2023
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	30/08/2023	30/08/2023
					CT2406	Implantable	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	30/08/2023	30/08/2023
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	30/08/2023	30/08/2023
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	30/08/2023	30/08/2023
					CT2407	Active implantable	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	30/08/2023	30/08/2023
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	30/08/2023	30/08/2023
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	30/08/2023	30/08/2023
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	30/08/2023	30/08/2023
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	30/08/2023	30/08/2023
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	30/08/2023	30/08/2023
					CT2727	Active implant devices/ancillaries	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	30/08/2023	30/08/2023
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	30/08/2023	30/08/2023
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	30/08/2023	30/08/2023
					CT2750	Implantable cardiac electrical stimulators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	30/08/2023	30/08/2023
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	30/08/2023	30/08/2023
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	30/08/2023	30/08/2023
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	30/08/2023	30/08/2023
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	30/08/2023	30/08/2023
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	30/08/2023	30/08/2023
					CT2985	Active	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	30/08/2023	30/08/2023
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	30/08/2023	30/08/2023
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	30/08/2023	30/08/2023
					CT3001	Implantable metal	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	30/08/2023	30/08/2023
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	30/08/2023	30/08/2023
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	30/08/2023	30/08/2023
					CT3002	Implantable synthetic polymer	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	30/08/2023	30/08/2023
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	30/08/2023	30/08/2023
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	30/08/2023	30/08/2023
					CT3006	Implantable inorganic material	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	30/08/2023	30/08/2023
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	30/08/2023	30/08/2023
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	30/08/2023	30/08/2023
					CT314	Electrically powered	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	30/08/2023	30/08/2023
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	30/08/2023	30/08/2023
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	30/08/2023	30/08/2023
					CT321	Long-term surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	30/08/2023	30/08/2023
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	30/08/2023	30/08/2023
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	30/08/2023	30/08/2023
					CT449	Generators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	30/08/2023	30/08/2023
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	30/08/2023	30/08/2023
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	30/08/2023	30/08/2023
					CT981	Single use	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	30/08/2023	30/08/2023
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	30/08/2023	30/08/2023
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	30/08/2023	30/08/2023
					CT983	Surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	30/08/2023	30/08/2023
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	30/08/2023	30/08/2023
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	30/08/2023	30/08/2023
14129	2023		37265	Dual-chamber implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E21	Procedural Complications	A16	Protective Measures Problem		
							E22	Investigations and Diagnostic Tests	A16	Protective Measures Problem		
							F26	No Health Consequences or Impact	A16	Protective Measures Problem		
14130	2023		47265	Dual-chamber implantable pacemaker, rate-responsive	CT141	Electrodes/Leads and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	31/08/2023	31/08/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	31/08/2023	31/08/2023
					CT145	Cardiovascular devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	31/08/2023	31/08/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	31/08/2023	31/08/2023
					CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	31/08/2023	31/08/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	31/08/2023	31/08/2023
					CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	31/08/2023	31/08/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	31/08/2023	31/08/2023
					CT233	Surgical	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	31/08/2023	31/08/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	31/08/2023	31/08/2023
					CT2406	Implantable	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	31/08/2023	31/08/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	31/08/2023	31/08/2023
					CT2407	Active implantable	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	31/08/2023	31/08/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	31/08/2023	31/08/2023
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	31/08/2023	31/08/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	31/08/2023	31/08/2023
					CT2727	Active implant devices/ancillaries	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	31/08/2023	31/08/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	31/08/2023	31/08/2023
					CT2750	Implantable cardiac electrical stimulators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	31/08/2023	31/08/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	31/08/2023	31/08/2023
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	31/08/2023	31/08/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	31/08/2023	31/08/2023
					CT2985	Active	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	31/08/2023	31/08/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	31/08/2023	31/08/2023
					CT3001	Implantable metal	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	31/08/2023	31/08/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	31/08/2023	31/08/2023

					CT3002	Implantable synthetic polymer	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	31/08/2023	31/08/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	31/08/2023	31/08/2023
					CT3006	Implantable inorganic material	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	31/08/2023	31/08/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	31/08/2023	31/08/2023
					CT314	Electrically powered	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	31/08/2023	31/08/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	31/08/2023	31/08/2023
					CT321	Long-term surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	31/08/2023	31/08/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	31/08/2023	31/08/2023
					CT449	Generators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	31/08/2023	31/08/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	31/08/2023	31/08/2023
					CT981	Single use	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	31/08/2023	31/08/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	31/08/2023	31/08/2023
					CT983	Surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	31/08/2023	31/08/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	31/08/2023	31/08/2023
14131	2023	All other reportable incidents	47265	Dual-chamber implantable pacemaker, rate-responsive	CT141	Electrodes/Leads and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A0723	Unexpected Mode Switch	25/08/2023	25/08/2023
							F1901	Additional Surgery	A0723	Unexpected Mode Switch	25/08/2023	25/08/2023
					CT145	Cardiovascular devices	E2403	No Clinical Signs, Symptoms or Conditions	A0723	Unexpected Mode Switch	25/08/2023	25/08/2023
							F1901	Additional Surgery	A0723	Unexpected Mode Switch	25/08/2023	25/08/2023
					CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A0723	Unexpected Mode Switch	25/08/2023	25/08/2023
							F1901	Additional Surgery	A0723	Unexpected Mode Switch	25/08/2023	25/08/2023
					CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A0723	Unexpected Mode Switch	25/08/2023	25/08/2023
							F1901	Additional Surgery	A0723	Unexpected Mode Switch	25/08/2023	25/08/2023
					CT233	Surgical	E2403	No Clinical Signs, Symptoms or Conditions	A0723	Unexpected Mode Switch	25/08/2023	25/08/2023
							F1901	Additional Surgery	A0723	Unexpected Mode Switch	25/08/2023	25/08/2023
					CT2406	Implantable	E2403	No Clinical Signs, Symptoms or Conditions	A0723	Unexpected Mode Switch	25/08/2023	25/08/2023
							F1901	Additional Surgery	A0723	Unexpected Mode Switch	25/08/2023	25/08/2023
					CT2407	Active implantable	E2403	No Clinical Signs, Symptoms or Conditions	A0723	Unexpected Mode Switch	25/08/2023	25/08/2023
							F1901	Additional Surgery	A0723	Unexpected Mode Switch	25/08/2023	25/08/2023
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A0723	Unexpected Mode Switch	25/08/2023	25/08/2023
							F1901	Additional Surgery	A0723	Unexpected Mode Switch	25/08/2023	25/08/2023
					CT2727	Active implant devices/ancillaries	E2403	No Clinical Signs, Symptoms or Conditions	A0723	Unexpected Mode Switch	25/08/2023	25/08/2023
							F1901	Additional Surgery	A0723	Unexpected Mode Switch	25/08/2023	25/08/2023
					CT2750	Implantable cardiac electrical stimulators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A0723	Unexpected Mode Switch	25/08/2023	25/08/2023
							F1901	Additional Surgery	A0723	Unexpected Mode Switch	25/08/2023	25/08/2023
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A0723	Unexpected Mode Switch	25/08/2023	25/08/2023
							F1901	Additional Surgery	A0723	Unexpected Mode Switch	25/08/2023	25/08/2023
					CT2985	Active	E2403	No Clinical Signs, Symptoms or Conditions	A0723	Unexpected Mode Switch	25/08/2023	25/08/2023
							F1901	Additional Surgery	A0723	Unexpected Mode Switch	25/08/2023	25/08/2023
					CT3001	Implantable metal	E2403	No Clinical Signs, Symptoms or Conditions	A0723	Unexpected Mode Switch	25/08/2023	25/08/2023
							F1901	Additional Surgery	A0723	Unexpected Mode Switch	25/08/2023	25/08/2023
					CT3002	Implantable synthetic polymer	E2403	No Clinical Signs, Symptoms or Conditions	A0723	Unexpected Mode Switch	25/08/2023	25/08/2023
							F1901	Additional Surgery	A0723	Unexpected Mode Switch	25/08/2023	25/08/2023
					CT3006	Implantable inorganic material	E2403	No Clinical Signs, Symptoms or Conditions	A0723	Unexpected Mode Switch	25/08/2023	25/08/2023
							F1901	Additional Surgery	A0723	Unexpected Mode Switch	25/08/2023	25/08/2023
					CT314	Electrically powered	E2403	No Clinical Signs, Symptoms or Conditions	A0723	Unexpected Mode Switch	25/08/2023	25/08/2023
							F1901	Additional Surgery	A0723	Unexpected Mode Switch	25/08/2023	25/08/2023
					CT321	Long-term surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A0723	Unexpected Mode Switch	25/08/2023	25/08/2023
							F1901	Additional Surgery	A0723	Unexpected Mode Switch	25/08/2023	25/08/2023
					CT449	Generators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A0723	Unexpected Mode Switch	25/08/2023	25/08/2023
							F1901	Additional Surgery	A0723	Unexpected Mode Switch	25/08/2023	25/08/2023
					CT981	Single use	E2403	No Clinical Signs, Symptoms or Conditions	A0723	Unexpected Mode Switch	25/08/2023	25/08/2023
							F1901	Additional Surgery	A0723	Unexpected Mode Switch	25/08/2023	25/08/2023
					CT983	Surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A0723	Unexpected Mode Switch	25/08/2023	25/08/2023
							F1901	Additional Surgery	A0723	Unexpected Mode Switch	25/08/2023	25/08/2023
14132	2023	All other reportable incidents	35223	Endocardial/interventricular septal pacing lead	CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A072202	Low Impedance		
							F1901	Additional Surgery	A072202	Low Impedance		
14133	2023		47265	Dual-chamber implantable pacemaker, rate-responsive	CT141	Electrodes/Leads and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	26/09/2023	26/09/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	26/09/2023	26/09/2023
					CT145	Cardiovascular devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	26/09/2023	26/09/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	26/09/2023	26/09/2023
					CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	26/09/2023	26/09/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	26/09/2023	26/09/2023
					CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	26/09/2023	26/09/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	26/09/2023	26/09/2023
					CT233	Surgical	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	26/09/2023	26/09/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	26/09/2023	26/09/2023
					CT2406	Implantable	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	26/09/2023	26/09/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	26/09/2023	26/09/2023
					CT2407	Active implantable	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	26/09/2023	26/09/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	26/09/2023	26/09/2023
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	26/09/2023	26/09/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	26/09/2023	26/09/2023
					CT2727	Active implant devices/ancillaries	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	26/09/2023	26/09/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	26/09/2023	26/09/2023
					CT2750	Implantable cardiac electrical stimulators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	26/09/2023	26/09/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	26/09/2023	26/09/2023
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	26/09/2023	26/09/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	26/09/2023	26/09/2023
					CT2985	Active	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	26/09/2023	26/09/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	26/09/2023	26/09/2023
					CT3001	Implantable metal	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	26/09/2023	26/09/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	26/09/2023	26/09/2023
					CT3002	Implantable synthetic polymer	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	26/09/2023	26/09/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	26/09/2023	26/09/2023
					CT3006	Implantable inorganic material	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	26/09/2023	26/09/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	26/09/2023	26/09/2023
					CT314	Electrically powered	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	26/09/2023	26/09/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	26/09/2023	26/09/2023

					CT321	Long-term surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	26/09/2023	26/09/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	26/09/2023	26/09/2023
					CT449	Generators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	26/09/2023	26/09/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	26/09/2023	26/09/2023
					CT981	Single use	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	26/09/2023	26/09/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	26/09/2023	26/09/2023
					CT983	Surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	26/09/2023	26/09/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	26/09/2023	26/09/2023
14134	2023	All other reportable incidents	47265	Dual-chamber implantable pacemaker, rate-responsive	CT141	Electrodes/Leads and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	19/07/2023	19/07/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	19/07/2023	19/07/2023
					CT145	Cardiovascular devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	19/07/2023	19/07/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	19/07/2023	19/07/2023
					CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	19/07/2023	19/07/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	19/07/2023	19/07/2023
					CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	19/07/2023	19/07/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	19/07/2023	19/07/2023
					CT233	Surgical	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	19/07/2023	19/07/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	19/07/2023	19/07/2023
					CT2406	Implantable	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	19/07/2023	19/07/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	19/07/2023	19/07/2023
					CT2407	Active implantable	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	19/07/2023	19/07/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	19/07/2023	19/07/2023
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	19/07/2023	19/07/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	19/07/2023	19/07/2023
					CT2727	Active implant devices/ancillaries	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	19/07/2023	19/07/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	19/07/2023	19/07/2023
					CT2750	Implantable cardiac electrical stimulators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	19/07/2023	19/07/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	19/07/2023	19/07/2023
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	19/07/2023	19/07/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	19/07/2023	19/07/2023
					CT2985	Active	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	19/07/2023	19/07/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	19/07/2023	19/07/2023
					CT3001	Implantable metal	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	19/07/2023	19/07/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	19/07/2023	19/07/2023
					CT3002	Implantable synthetic polymer	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	19/07/2023	19/07/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	19/07/2023	19/07/2023
					CT3006	Implantable inorganic material	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	19/07/2023	19/07/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	19/07/2023	19/07/2023
					CT314	Electrically powered	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	19/07/2023	19/07/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	19/07/2023	19/07/2023
					CT321	Long-term surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	19/07/2023	19/07/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	19/07/2023	19/07/2023
					CT449	Generators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	19/07/2023	19/07/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	19/07/2023	19/07/2023
					CT981	Single use	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	19/07/2023	19/07/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	19/07/2023	19/07/2023
					CT983	Surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	19/07/2023	19/07/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	19/07/2023	19/07/2023
14135	2023	All other reportable incidents	47265	Dual-chamber implantable pacemaker, rate-responsive	CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	16/06/2023	16/06/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	16/06/2023	16/06/2023
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	16/06/2023	16/06/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	16/06/2023	16/06/2023
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	16/06/2023	16/06/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	16/06/2023	16/06/2023
14136	2023	All other reportable incidents	47265	Dual-chamber implantable pacemaker, rate-responsive	CT141	Electrodes/Leads and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070909	Oversensing		
							F26	No Health Consequences or Impact	A070909	Oversensing		
					CT145	Cardiovascular devices	E2403	No Clinical Signs, Symptoms or Conditions	A070909	Oversensing		
							F26	No Health Consequences or Impact	A070909	Oversensing		
					CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070909	Oversensing		
							F26	No Health Consequences or Impact	A070909	Oversensing		
					CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A070909	Oversensing		
							F26	No Health Consequences or Impact	A070909	Oversensing		
					CT233	Surgical	E2403	No Clinical Signs, Symptoms or Conditions	A070909	Oversensing		
							F26	No Health Consequences or Impact	A070909	Oversensing		
					CT2406	Implantable	E2403	No Clinical Signs, Symptoms or Conditions	A070909	Oversensing		
							F26	No Health Consequences or Impact	A070909	Oversensing		
					CT2407	Active implantable	E2403	No Clinical Signs, Symptoms or Conditions	A070909	Oversensing		
							F26	No Health Consequences or Impact	A070909	Oversensing		
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A070909	Oversensing		
							F26	No Health Consequences or Impact	A070909	Oversensing		
					CT2727	Active implant devices/ancillaries	E2403	No Clinical Signs, Symptoms or Conditions	A070909	Oversensing		
							F26	No Health Consequences or Impact	A070909	Oversensing		
					CT2750	Implantable cardiac electrical stimulators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070909	Oversensing		
							F26	No Health Consequences or Impact	A070909	Oversensing		
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A070909	Oversensing		
							F26	No Health Consequences or Impact	A070909	Oversensing		
					CT2985	Active	E2403	No Clinical Signs, Symptoms or Conditions	A070909	Oversensing		
							F26	No Health Consequences or Impact	A070909	Oversensing		
					CT3001	Implantable metal	E2403	No Clinical Signs, Symptoms or Conditions	A070909	Oversensing		
							F26	No Health Consequences or Impact	A070909	Oversensing		
					CT3002	Implantable synthetic polymer	E2403	No Clinical Signs, Symptoms or Conditions	A070909	Oversensing		
							F26	No Health Consequences or Impact	A070909	Oversensing		
					CT3006	Implantable inorganic material	E2403	No Clinical Signs, Symptoms or Conditions	A070909	Oversensing		
							F26	No Health Consequences or Impact	A070909	Oversensing		
					CT314	Electrically powered	E2403	No Clinical Signs, Symptoms or Conditions	A070909	Oversensing		
							F26	No Health Consequences or Impact	A070909	Oversensing		
					CT321	Long-term surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A070909	Oversensing		
							F26	No Health Consequences or Impact	A070909	Oversensing		

					CT449	Generators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070909	Oversensing		
							F26	No Health Consequences or Impact	A070909	Oversensing		
					CT981	Single use	E2403	No Clinical Signs, Symptoms or Conditions	A070909	Oversensing		
							F26	No Health Consequences or Impact	A070909	Oversensing		
					CT983	Surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A070909	Oversensing		
							F26	No Health Consequences or Impact	A070909	Oversensing		
14137	2023		47265	Dual-chamber implantable pacemaker, rate-responsive	CT141	Electrodes/Leads and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	23/03/2023	23/03/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	23/03/2023	23/03/2023
					CT145	Cardiovascular devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	23/03/2023	23/03/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	23/03/2023	23/03/2023
					CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	23/03/2023	23/03/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	23/03/2023	23/03/2023
					CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	23/03/2023	23/03/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	23/03/2023	23/03/2023
					CT233	Surgical	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	23/03/2023	23/03/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	23/03/2023	23/03/2023
					CT2406	Implantable	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	23/03/2023	23/03/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	23/03/2023	23/03/2023
					CT2407	Active implantable	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	23/03/2023	23/03/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	23/03/2023	23/03/2023
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	23/03/2023	23/03/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	23/03/2023	23/03/2023
					CT2727	Active implant devices/ancillaries	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	23/03/2023	23/03/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	23/03/2023	23/03/2023
					CT2750	Implantable cardiac electrical stimulators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	23/03/2023	23/03/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	23/03/2023	23/03/2023
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	23/03/2023	23/03/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	23/03/2023	23/03/2023
					CT2985	Active	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	23/03/2023	23/03/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	23/03/2023	23/03/2023
					CT3001	Implantable metal	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	23/03/2023	23/03/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	23/03/2023	23/03/2023
					CT3002	Implantable synthetic polymer	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	23/03/2023	23/03/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	23/03/2023	23/03/2023
					CT3006	Implantable inorganic material	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	23/03/2023	23/03/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	23/03/2023	23/03/2023
					CT314	Electrically powered	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	23/03/2023	23/03/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	23/03/2023	23/03/2023
					CT321	Long-term surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	23/03/2023	23/03/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	23/03/2023	23/03/2023
					CT449	Generators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	23/03/2023	23/03/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	23/03/2023	23/03/2023
					CT981	Single use	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	23/03/2023	23/03/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	23/03/2023	23/03/2023
					CT983	Surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	23/03/2023	23/03/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	23/03/2023	23/03/2023
14138	2023	All other reportable incidents	47270	Cardiac resynchronization therapy implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
14139	2023		47265	Dual-chamber implantable pacemaker, rate-responsive	CT141	Electrodes/Leads and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	20/09/2023	20/09/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	20/09/2023	20/09/2023
					CT145	Cardiovascular devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	20/09/2023	20/09/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	20/09/2023	20/09/2023
					CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	20/09/2023	20/09/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	20/09/2023	20/09/2023
					CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	20/09/2023	20/09/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	20/09/2023	20/09/2023
					CT233	Surgical	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	20/09/2023	20/09/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	20/09/2023	20/09/2023
					CT2406	Implantable	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	20/09/2023	20/09/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	20/09/2023	20/09/2023
					CT2407	Active implantable	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	20/09/2023	20/09/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	20/09/2023	20/09/2023
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	20/09/2023	20/09/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	20/09/2023	20/09/2023
					CT2727	Active implant devices/ancillaries	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	20/09/2023	20/09/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	20/09/2023	20/09/2023
					CT2750	Implantable cardiac electrical stimulators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	20/09/2023	20/09/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	20/09/2023	20/09/2023
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	20/09/2023	20/09/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	20/09/2023	20/09/2023
					CT2985	Active	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	20/09/2023	20/09/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	20/09/2023	20/09/2023
					CT3001	Implantable metal	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	20/09/2023	20/09/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	20/09/2023	20/09/2023
					CT3002	Implantable synthetic polymer	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	20/09/2023	20/09/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	20/09/2023	20/09/2023
					CT3006	Implantable inorganic material	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	20/09/2023	20/09/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	20/09/2023	20/09/2023
					CT314	Electrically powered	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	20/09/2023	20/09/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	20/09/2023	20/09/2023
					CT321	Long-term surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	20/09/2023	20/09/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	20/09/2023	20/09/2023
					CT449	Generators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	20/09/2023	20/09/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	20/09/2023	20/09/2023
					CT981	Single use	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	20/09/2023	20/09/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	20/09/2023	20/09/2023
					CT983	Surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	20/09/2023	20/09/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	20/09/2023	20/09/2023

14140	2023	All other reportable incidents	47205	Cardiac pulse generator programmer	CT1684	Cardiac pacemakers and associated devices	E0601	Arrhythmia	A0719	Unexpected Shutdown		
							F12	Serious Injury/ Illness/ Impairment	A0719	Unexpected Shutdown		
14141	2023		47265	Dual-chamber implantable pacemaker, rate-responsive	CT141	Electrodes/Leads and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	22/09/2023	22/09/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	22/09/2023	22/09/2023
					CT145	Cardiovascular devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	22/09/2023	22/09/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	22/09/2023	22/09/2023
					CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	22/09/2023	22/09/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	22/09/2023	22/09/2023
					CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	22/09/2023	22/09/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	22/09/2023	22/09/2023
					CT233	Surgical	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	22/09/2023	22/09/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	22/09/2023	22/09/2023
					CT2406	Implantable	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	22/09/2023	22/09/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	22/09/2023	22/09/2023
					CT2407	Active implantable	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	22/09/2023	22/09/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	22/09/2023	22/09/2023
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	22/09/2023	22/09/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	22/09/2023	22/09/2023
					CT2727	Active implant devices/ancillaries	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	22/09/2023	22/09/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	22/09/2023	22/09/2023
					CT2750	Implantable cardiac electrical stimulators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	22/09/2023	22/09/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	22/09/2023	22/09/2023
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	22/09/2023	22/09/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	22/09/2023	22/09/2023
					CT2985	Active	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	22/09/2023	22/09/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	22/09/2023	22/09/2023
					CT3001	Implantable metal	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	22/09/2023	22/09/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	22/09/2023	22/09/2023
					CT3002	Implantable synthetic polymer	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	22/09/2023	22/09/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	22/09/2023	22/09/2023
					CT3006	Implantable inorganic material	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	22/09/2023	22/09/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	22/09/2023	22/09/2023
					CT314	Electrically powered	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	22/09/2023	22/09/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	22/09/2023	22/09/2023
					CT321	Long-term surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	22/09/2023	22/09/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	22/09/2023	22/09/2023
					CT449	Generators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	22/09/2023	22/09/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	22/09/2023	22/09/2023
					CT981	Single use	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	22/09/2023	22/09/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	22/09/2023	22/09/2023
					CT983	Surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	22/09/2023	22/09/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	22/09/2023	22/09/2023
14142	2023	All other reportable incidents	36241	Pacemaker/defibrillator lead	CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070909	Oversensing		
									A072201	High Impedance		
							F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A070909	Oversensing		
									A072201	High Impedance		
14143	2023	All other reportable incidents	47265	Dual-chamber implantable pacemaker, rate-responsive	CT141	Electrodes/Leads and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A0908	Incorrect, Inadequate or Imprecise Result or Readings	30/06/2023	30/06/2023
							F08	Hospitalization or Prolonged Hospitalization	A0908	Incorrect, Inadequate or Imprecise Result or Readings	30/06/2023	30/06/2023
							F1905	Device Revision or Replacement	A0908	Incorrect, Inadequate or Imprecise Result or Readings	30/06/2023	30/06/2023
					CT145	Cardiovascular devices	E2403	No Clinical Signs, Symptoms or Conditions	A0908	Incorrect, Inadequate or Imprecise Result or Readings	30/06/2023	30/06/2023
							F08	Hospitalization or Prolonged Hospitalization	A0908	Incorrect, Inadequate or Imprecise Result or Readings	30/06/2023	30/06/2023
							F1905	Device Revision or Replacement	A0908	Incorrect, Inadequate or Imprecise Result or Readings	30/06/2023	30/06/2023
					CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A0908	Incorrect, Inadequate or Imprecise Result or Readings	30/06/2023	30/06/2023
							F08	Hospitalization or Prolonged Hospitalization	A0908	Incorrect, Inadequate or Imprecise Result or Readings	30/06/2023	30/06/2023
							F1905	Device Revision or Replacement	A0908	Incorrect, Inadequate or Imprecise Result or Readings	30/06/2023	30/06/2023
					CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A0908	Incorrect, Inadequate or Imprecise Result or Readings	30/06/2023	30/06/2023
							F08	Hospitalization or Prolonged Hospitalization	A0908	Incorrect, Inadequate or Imprecise Result or Readings	30/06/2023	30/06/2023
							F1905	Device Revision or Replacement	A0908	Incorrect, Inadequate or Imprecise Result or Readings	30/06/2023	30/06/2023
					CT233	Surgical	E2403	No Clinical Signs, Symptoms or Conditions	A0908	Incorrect, Inadequate or Imprecise Result or Readings	30/06/2023	30/06/2023
							F08	Hospitalization or Prolonged Hospitalization	A0908	Incorrect, Inadequate or Imprecise Result or Readings	30/06/2023	30/06/2023
							F1905	Device Revision or Replacement	A0908	Incorrect, Inadequate or Imprecise Result or Readings	30/06/2023	30/06/2023
					CT2406	Implantable	E2403	No Clinical Signs, Symptoms or Conditions	A0908	Incorrect, Inadequate or Imprecise Result or Readings	30/06/2023	30/06/2023
							F08	Hospitalization or Prolonged Hospitalization	A0908	Incorrect, Inadequate or Imprecise Result or Readings	30/06/2023	30/06/2023
							F1905	Device Revision or Replacement	A0908	Incorrect, Inadequate or Imprecise Result or Readings	30/06/2023	30/06/2023
					CT2407	Active implantable	E2403	No Clinical Signs, Symptoms or Conditions	A0908	Incorrect, Inadequate or Imprecise Result or Readings	30/06/2023	30/06/2023
							F08	Hospitalization or Prolonged Hospitalization	A0908	Incorrect, Inadequate or Imprecise Result or Readings	30/06/2023	30/06/2023
							F1905	Device Revision or Replacement	A0908	Incorrect, Inadequate or Imprecise Result or Readings	30/06/2023	30/06/2023
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A0908	Incorrect, Inadequate or Imprecise Result or Readings	30/06/2023	30/06/2023
							F08	Hospitalization or Prolonged Hospitalization	A0908	Incorrect, Inadequate or Imprecise Result or Readings	30/06/2023	30/06/2023

							F1905	Device Revision or Replacement	A0908	Incorrect, Inadequate or Imprecise Result or Readings	30/06/2023	30/06/2023
					CT2727	Active implant devices/ancillaries	E2403	No Clinical Signs, Symptoms or Conditions	A0908	Incorrect, Inadequate or Imprecise Result or Readings	30/06/2023	30/06/2023
							F08	Hospitalization or Prolonged Hospitalization	A0908	Incorrect, Inadequate or Imprecise Result or Readings	30/06/2023	30/06/2023
							F1905	Device Revision or Replacement	A0908	Incorrect, Inadequate or Imprecise Result or Readings	30/06/2023	30/06/2023
					CT2750	Implantable cardiac electrical stimulators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A0908	Incorrect, Inadequate or Imprecise Result or Readings	30/06/2023	30/06/2023
							F08	Hospitalization or Prolonged Hospitalization	A0908	Incorrect, Inadequate or Imprecise Result or Readings	30/06/2023	30/06/2023
							F1905	Device Revision or Replacement	A0908	Incorrect, Inadequate or Imprecise Result or Readings	30/06/2023	30/06/2023
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A0908	Incorrect, Inadequate or Imprecise Result or Readings	30/06/2023	30/06/2023
							F08	Hospitalization or Prolonged Hospitalization	A0908	Incorrect, Inadequate or Imprecise Result or Readings	30/06/2023	30/06/2023
							F1905	Device Revision or Replacement	A0908	Incorrect, Inadequate or Imprecise Result or Readings	30/06/2023	30/06/2023
					CT2985	Active	E2403	No Clinical Signs, Symptoms or Conditions	A0908	Incorrect, Inadequate or Imprecise Result or Readings	30/06/2023	30/06/2023
							F08	Hospitalization or Prolonged Hospitalization	A0908	Incorrect, Inadequate or Imprecise Result or Readings	30/06/2023	30/06/2023
							F1905	Device Revision or Replacement	A0908	Incorrect, Inadequate or Imprecise Result or Readings	30/06/2023	30/06/2023
					CT3001	Implantable metal	E2403	No Clinical Signs, Symptoms or Conditions	A0908	Incorrect, Inadequate or Imprecise Result or Readings	30/06/2023	30/06/2023
							F08	Hospitalization or Prolonged Hospitalization	A0908	Incorrect, Inadequate or Imprecise Result or Readings	30/06/2023	30/06/2023
							F1905	Device Revision or Replacement	A0908	Incorrect, Inadequate or Imprecise Result or Readings	30/06/2023	30/06/2023
					CT3002	Implantable synthetic polymer	E2403	No Clinical Signs, Symptoms or Conditions	A0908	Incorrect, Inadequate or Imprecise Result or Readings	30/06/2023	30/06/2023
							F08	Hospitalization or Prolonged Hospitalization	A0908	Incorrect, Inadequate or Imprecise Result or Readings	30/06/2023	30/06/2023
							F1905	Device Revision or Replacement	A0908	Incorrect, Inadequate or Imprecise Result or Readings	30/06/2023	30/06/2023
					CT3006	Implantable inorganic material	E2403	No Clinical Signs, Symptoms or Conditions	A0908	Incorrect, Inadequate or Imprecise Result or Readings	30/06/2023	30/06/2023
							F08	Hospitalization or Prolonged Hospitalization	A0908	Incorrect, Inadequate or Imprecise Result or Readings	30/06/2023	30/06/2023
							F1905	Device Revision or Replacement	A0908	Incorrect, Inadequate or Imprecise Result or Readings	30/06/2023	30/06/2023
					CT314	Electrically powered	E2403	No Clinical Signs, Symptoms or Conditions	A0908	Incorrect, Inadequate or Imprecise Result or Readings	30/06/2023	30/06/2023
							F08	Hospitalization or Prolonged Hospitalization	A0908	Incorrect, Inadequate or Imprecise Result or Readings	30/06/2023	30/06/2023
							F1905	Device Revision or Replacement	A0908	Incorrect, Inadequate or Imprecise Result or Readings	30/06/2023	30/06/2023
					CT321	Long-term surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A0908	Incorrect, Inadequate or Imprecise Result or Readings	30/06/2023	30/06/2023
							F08	Hospitalization or Prolonged Hospitalization	A0908	Incorrect, Inadequate or Imprecise Result or Readings	30/06/2023	30/06/2023
							F1905	Device Revision or Replacement	A0908	Incorrect, Inadequate or Imprecise Result or Readings	30/06/2023	30/06/2023
					CT449	Generators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A0908	Incorrect, Inadequate or Imprecise Result or Readings	30/06/2023	30/06/2023
							F08	Hospitalization or Prolonged Hospitalization	A0908	Incorrect, Inadequate or Imprecise Result or Readings	30/06/2023	30/06/2023
							F1905	Device Revision or Replacement	A0908	Incorrect, Inadequate or Imprecise Result or Readings	30/06/2023	30/06/2023
					CT981	Single use	E2403	No Clinical Signs, Symptoms or Conditions	A0908	Incorrect, Inadequate or Imprecise Result or Readings	30/06/2023	30/06/2023
							F08	Hospitalization or Prolonged Hospitalization	A0908	Incorrect, Inadequate or Imprecise Result or Readings	30/06/2023	30/06/2023
							F1905	Device Revision or Replacement	A0908	Incorrect, Inadequate or Imprecise Result or Readings	30/06/2023	30/06/2023
					CT983	Surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A0908	Incorrect, Inadequate or Imprecise Result or Readings	30/06/2023	30/06/2023
							F08	Hospitalization or Prolonged Hospitalization	A0908	Incorrect, Inadequate or Imprecise Result or Readings	30/06/2023	30/06/2023
							F1905	Device Revision or Replacement	A0908	Incorrect, Inadequate or Imprecise Result or Readings	30/06/2023	30/06/2023
14144	2023	All other reportable incidents	47265	Dual-chamber implantable pacemaker, rate-responsive	CT141	Electrodes/Leads and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A0908	Incorrect, Inadequate or Imprecise Result or Readings	01/08/2023	01/08/2023
							F08	Hospitalization or Prolonged Hospitalization	A0908	Incorrect, Inadequate or Imprecise Result or Readings	01/08/2023	01/08/2023
							F1905	Device Revision or Replacement	A0908	Incorrect, Inadequate or Imprecise Result or Readings	01/08/2023	01/08/2023
					CT145	Cardiovascular devices	E2403	No Clinical Signs, Symptoms or Conditions	A0908	Incorrect, Inadequate or Imprecise Result or Readings	01/08/2023	01/08/2023
							F08	Hospitalization or Prolonged Hospitalization	A0908	Incorrect, Inadequate or Imprecise Result or Readings	01/08/2023	01/08/2023
							F1905	Device Revision or Replacement	A0908	Incorrect, Inadequate or Imprecise Result or Readings	01/08/2023	01/08/2023
					CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A0908	Incorrect, Inadequate or Imprecise Result or Readings	01/08/2023	01/08/2023
							F08	Hospitalization or Prolonged Hospitalization	A0908	Incorrect, Inadequate or Imprecise Result or Readings	01/08/2023	01/08/2023
							F1905	Device Revision or Replacement	A0908	Incorrect, Inadequate or Imprecise Result or Readings	01/08/2023	01/08/2023
					CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A0908	Incorrect, Inadequate or Imprecise Result or Readings	01/08/2023	01/08/2023
							F08	Hospitalization or Prolonged Hospitalization	A0908	Incorrect, Inadequate or Imprecise Result or Readings	01/08/2023	01/08/2023
							F1905	Device Revision or Replacement	A0908	Incorrect, Inadequate or Imprecise Result or Readings	01/08/2023	01/08/2023
					CT233	Surgical	E2403	No Clinical Signs, Symptoms or Conditions	A0908	Incorrect, Inadequate or Imprecise Result or Readings	01/08/2023	01/08/2023

							F08	Hospitalization or Prolonged Hospitalization	A0908	Incorrect, Inadequate or Imprecise Result or Readings	01/08/2023	01/08/2023
							F1905	Device Revision or Replacement	A0908	Incorrect, Inadequate or Imprecise Result or Readings	01/08/2023	01/08/2023
					CT2406	Implantable	E2403	No Clinical Signs, Symptoms or Conditions	A0908	Incorrect, Inadequate or Imprecise Result or Readings	01/08/2023	01/08/2023
							F08	Hospitalization or Prolonged Hospitalization	A0908	Incorrect, Inadequate or Imprecise Result or Readings	01/08/2023	01/08/2023
							F1905	Device Revision or Replacement	A0908	Incorrect, Inadequate or Imprecise Result or Readings	01/08/2023	01/08/2023
					CT2407	Active implantable	E2403	No Clinical Signs, Symptoms or Conditions	A0908	Incorrect, Inadequate or Imprecise Result or Readings	01/08/2023	01/08/2023
							F08	Hospitalization or Prolonged Hospitalization	A0908	Incorrect, Inadequate or Imprecise Result or Readings	01/08/2023	01/08/2023
							F1905	Device Revision or Replacement	A0908	Incorrect, Inadequate or Imprecise Result or Readings	01/08/2023	01/08/2023
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A0908	Incorrect, Inadequate or Imprecise Result or Readings	01/08/2023	01/08/2023
							F08	Hospitalization or Prolonged Hospitalization	A0908	Incorrect, Inadequate or Imprecise Result or Readings	01/08/2023	01/08/2023
							F1905	Device Revision or Replacement	A0908	Incorrect, Inadequate or Imprecise Result or Readings	01/08/2023	01/08/2023
					CT2727	Active implant devices/ancillaries	E2403	No Clinical Signs, Symptoms or Conditions	A0908	Incorrect, Inadequate or Imprecise Result or Readings	01/08/2023	01/08/2023
							F08	Hospitalization or Prolonged Hospitalization	A0908	Incorrect, Inadequate or Imprecise Result or Readings	01/08/2023	01/08/2023
							F1905	Device Revision or Replacement	A0908	Incorrect, Inadequate or Imprecise Result or Readings	01/08/2023	01/08/2023
					CT2750	Implantable cardiac electrical stimulators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A0908	Incorrect, Inadequate or Imprecise Result or Readings	01/08/2023	01/08/2023
							F08	Hospitalization or Prolonged Hospitalization	A0908	Incorrect, Inadequate or Imprecise Result or Readings	01/08/2023	01/08/2023
							F1905	Device Revision or Replacement	A0908	Incorrect, Inadequate or Imprecise Result or Readings	01/08/2023	01/08/2023
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A0908	Incorrect, Inadequate or Imprecise Result or Readings	01/08/2023	01/08/2023
							F08	Hospitalization or Prolonged Hospitalization	A0908	Incorrect, Inadequate or Imprecise Result or Readings	01/08/2023	01/08/2023
							F1905	Device Revision or Replacement	A0908	Incorrect, Inadequate or Imprecise Result or Readings	01/08/2023	01/08/2023
					CT2985	Active	E2403	No Clinical Signs, Symptoms or Conditions	A0908	Incorrect, Inadequate or Imprecise Result or Readings	01/08/2023	01/08/2023
							F08	Hospitalization or Prolonged Hospitalization	A0908	Incorrect, Inadequate or Imprecise Result or Readings	01/08/2023	01/08/2023
							F1905	Device Revision or Replacement	A0908	Incorrect, Inadequate or Imprecise Result or Readings	01/08/2023	01/08/2023
					CT3001	Implantable metal	E2403	No Clinical Signs, Symptoms or Conditions	A0908	Incorrect, Inadequate or Imprecise Result or Readings	01/08/2023	01/08/2023
							F08	Hospitalization or Prolonged Hospitalization	A0908	Incorrect, Inadequate or Imprecise Result or Readings	01/08/2023	01/08/2023
							F1905	Device Revision or Replacement	A0908	Incorrect, Inadequate or Imprecise Result or Readings	01/08/2023	01/08/2023
					CT3002	Implantable synthetic polymer	E2403	No Clinical Signs, Symptoms or Conditions	A0908	Incorrect, Inadequate or Imprecise Result or Readings	01/08/2023	01/08/2023
							F08	Hospitalization or Prolonged Hospitalization	A0908	Incorrect, Inadequate or Imprecise Result or Readings	01/08/2023	01/08/2023
							F1905	Device Revision or Replacement	A0908	Incorrect, Inadequate or Imprecise Result or Readings	01/08/2023	01/08/2023
					CT3006	Implantable inorganic material	E2403	No Clinical Signs, Symptoms or Conditions	A0908	Incorrect, Inadequate or Imprecise Result or Readings	01/08/2023	01/08/2023
							F08	Hospitalization or Prolonged Hospitalization	A0908	Incorrect, Inadequate or Imprecise Result or Readings	01/08/2023	01/08/2023
							F1905	Device Revision or Replacement	A0908	Incorrect, Inadequate or Imprecise Result or Readings	01/08/2023	01/08/2023
					CT314	Electrically powered	E2403	No Clinical Signs, Symptoms or Conditions	A0908	Incorrect, Inadequate or Imprecise Result or Readings	01/08/2023	01/08/2023
							F08	Hospitalization or Prolonged Hospitalization	A0908	Incorrect, Inadequate or Imprecise Result or Readings	01/08/2023	01/08/2023
							F1905	Device Revision or Replacement	A0908	Incorrect, Inadequate or Imprecise Result or Readings	01/08/2023	01/08/2023
					CT321	Long-term surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A0908	Incorrect, Inadequate or Imprecise Result or Readings	01/08/2023	01/08/2023
							F08	Hospitalization or Prolonged Hospitalization	A0908	Incorrect, Inadequate or Imprecise Result or Readings	01/08/2023	01/08/2023
							F1905	Device Revision or Replacement	A0908	Incorrect, Inadequate or Imprecise Result or Readings	01/08/2023	01/08/2023
					CT449	Generators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A0908	Incorrect, Inadequate or Imprecise Result or Readings	01/08/2023	01/08/2023
							F08	Hospitalization or Prolonged Hospitalization	A0908	Incorrect, Inadequate or Imprecise Result or Readings	01/08/2023	01/08/2023
							F1905	Device Revision or Replacement	A0908	Incorrect, Inadequate or Imprecise Result or Readings	01/08/2023	01/08/2023
					CT981	Single use	E2403	No Clinical Signs, Symptoms or Conditions	A0908	Incorrect, Inadequate or Imprecise Result or Readings	01/08/2023	01/08/2023
							F08	Hospitalization or Prolonged Hospitalization	A0908	Incorrect, Inadequate or Imprecise Result or Readings	01/08/2023	01/08/2023
							F1905	Device Revision or Replacement	A0908	Incorrect, Inadequate or Imprecise Result or Readings	01/08/2023	01/08/2023
					CT983	Surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A0908	Incorrect, Inadequate or Imprecise Result or Readings	01/08/2023	01/08/2023
							F08	Hospitalization or Prolonged Hospitalization	A0908	Incorrect, Inadequate or Imprecise Result or Readings	01/08/2023	01/08/2023
							F1905	Device Revision or Replacement	A0908	Incorrect, Inadequate or Imprecise Result or Readings	01/08/2023	01/08/2023
14145	2023	All other reportable incidents	35852	Single-chamber implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E060101	Asystole	A070909	Oversensing		
									A0712	Pacing Problem		
									A072201	High Impedance		
									A090801	Signal Artifact/Noise		
							E2007	Fall	A070909	Oversensing		

								A0712	Pacing Problem		
								A072201	High Impedance		
								A090801	Signal Artifact/Noise		
						F08	Hospitalization or Prolonged Hospitalization	A070909	Oversensing		
								A0712	Pacing Problem		
								A072201	High Impedance		
								A090801	Signal Artifact/Noise		
						F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A070909	Oversensing		
								A0712	Pacing Problem		
								A072201	High Impedance		
								A090801	Signal Artifact/Noise		
14146	2023	All other reportable incidents	47265	Dual-chamber implantable pacemaker, rate-responsive	CT141	Electrodes/Leads and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070102	High Capture Threshold	28/06/2023	28/06/2023
									A0705	Battery Problem	28/06/2023	28/06/2023
									A070504	Premature Discharge of Battery	28/06/2023	28/06/2023
							F08	Hospitalization or Prolonged Hospitalization	A070102	High Capture Threshold	28/06/2023	28/06/2023
									A0705	Battery Problem	28/06/2023	28/06/2023
									A070504	Premature Discharge of Battery	28/06/2023	28/06/2023
							F1905	Device Revision or Replacement	A070102	High Capture Threshold	28/06/2023	28/06/2023
									A0705	Battery Problem	28/06/2023	28/06/2023
									A070504	Premature Discharge of Battery	28/06/2023	28/06/2023
					CT145	Cardiovascular devices	E2403	No Clinical Signs, Symptoms or Conditions	A070102	High Capture Threshold	28/06/2023	28/06/2023
									A0705	Battery Problem	28/06/2023	28/06/2023
									A070504	Premature Discharge of Battery	28/06/2023	28/06/2023
							F08	Hospitalization or Prolonged Hospitalization	A070102	High Capture Threshold	28/06/2023	28/06/2023
									A0705	Battery Problem	28/06/2023	28/06/2023
									A070504	Premature Discharge of Battery	28/06/2023	28/06/2023
							F1905	Device Revision or Replacement	A070102	High Capture Threshold	28/06/2023	28/06/2023
									A0705	Battery Problem	28/06/2023	28/06/2023
									A070504	Premature Discharge of Battery	28/06/2023	28/06/2023
					CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070102	High Capture Threshold	28/06/2023	28/06/2023
									A0705	Battery Problem	28/06/2023	28/06/2023
									A070504	Premature Discharge of Battery	28/06/2023	28/06/2023
							F08	Hospitalization or Prolonged Hospitalization	A070102	High Capture Threshold	28/06/2023	28/06/2023
									A0705	Battery Problem	28/06/2023	28/06/2023
									A070504	Premature Discharge of Battery	28/06/2023	28/06/2023
							F1905	Device Revision or Replacement	A070102	High Capture Threshold	28/06/2023	28/06/2023
									A0705	Battery Problem	28/06/2023	28/06/2023
									A070504	Premature Discharge of Battery	28/06/2023	28/06/2023
					CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A070102	High Capture Threshold	28/06/2023	28/06/2023
									A0705	Battery Problem	28/06/2023	28/06/2023
									A070504	Premature Discharge of Battery	28/06/2023	28/06/2023
							F08	Hospitalization or Prolonged Hospitalization	A070102	High Capture Threshold	28/06/2023	28/06/2023
									A0705	Battery Problem	28/06/2023	28/06/2023
									A070504	Premature Discharge of Battery	28/06/2023	28/06/2023
							F1905	Device Revision or Replacement	A070102	High Capture Threshold	28/06/2023	28/06/2023
									A0705	Battery Problem	28/06/2023	28/06/2023
									A070504	Premature Discharge of Battery	28/06/2023	28/06/2023
					CT233	Surgical	E2403	No Clinical Signs, Symptoms or Conditions	A070102	High Capture Threshold	28/06/2023	28/06/2023
									A0705	Battery Problem	28/06/2023	28/06/2023
									A070504	Premature Discharge of Battery	28/06/2023	28/06/2023
							F08	Hospitalization or Prolonged Hospitalization	A070102	High Capture Threshold	28/06/2023	28/06/2023
									A0705	Battery Problem	28/06/2023	28/06/2023
									A070504	Premature Discharge of Battery	28/06/2023	28/06/2023
							F1905	Device Revision or Replacement	A070102	High Capture Threshold	28/06/2023	28/06/2023
									A0705	Battery Problem	28/06/2023	28/06/2023
									A070504	Premature Discharge of Battery	28/06/2023	28/06/2023
					CT2406	Implantable	E2403	No Clinical Signs, Symptoms or Conditions	A070102	High Capture Threshold	28/06/2023	28/06/2023
									A0705	Battery Problem	28/06/2023	28/06/2023
									A070504	Premature Discharge of Battery	28/06/2023	28/06/2023
							F08	Hospitalization or Prolonged Hospitalization	A070102	High Capture Threshold	28/06/2023	28/06/2023
									A0705	Battery Problem	28/06/2023	28/06/2023
									A070504	Premature Discharge of Battery	28/06/2023	28/06/2023
							F1905	Device Revision or Replacement	A070102	High Capture Threshold	28/06/2023	28/06/2023
									A0705	Battery Problem	28/06/2023	28/06/2023
									A070504	Premature Discharge of Battery	28/06/2023	28/06/2023
					CT2407	Active implantable	E2403	No Clinical Signs, Symptoms or Conditions	A070102	High Capture Threshold	28/06/2023	28/06/2023
									A0705	Battery Problem	28/06/2023	28/06/2023
									A070504	Premature Discharge of Battery	28/06/2023	28/06/2023
							F08	Hospitalization or Prolonged Hospitalization	A070102	High Capture Threshold	28/06/2023	28/06/2023
									A0705	Battery Problem	28/06/2023	28/06/2023
									A070504	Premature Discharge of Battery	28/06/2023	28/06/2023
							F1905	Device Revision or Replacement	A070102	High Capture Threshold	28/06/2023	28/06/2023
									A0705	Battery Problem	28/06/2023	28/06/2023
									A070504	Premature Discharge of Battery	28/06/2023	28/06/2023
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A070102	High Capture Threshold	28/06/2023	28/06/2023
									A0705	Battery Problem	28/06/2023	28/06/2023
									A070504	Premature Discharge of Battery	28/06/2023	28/06/2023
							F08	Hospitalization or Prolonged Hospitalization	A070102	High Capture Threshold	28/06/2023	28/06/2023
									A0705	Battery Problem	28/06/2023	28/06/2023
									A070504	Premature Discharge of Battery	28/06/2023	28/06/2023
							F1905	Device Revision or Replacement	A070102	High Capture Threshold	28/06/2023	28/06/2023
									A0705	Battery Problem	28/06/2023	28/06/2023
									A070504	Premature Discharge of Battery	28/06/2023	28/06/2023
					CT2727	Active implant devices/ancillaries	E2403	No Clinical Signs, Symptoms or Conditions	A070102	High Capture Threshold	28/06/2023	28/06/2023
									A0705	Battery Problem	28/06/2023	28/06/2023
									A070504	Premature Discharge of Battery	28/06/2023	28/06/2023
							F08	Hospitalization or Prolonged Hospitalization	A070102	High Capture Threshold	28/06/2023	28/06/2023
									A0705	Battery Problem	28/06/2023	28/06/2023

								A070504	Premature Discharge of Battery	28/06/2023	28/06/2023
						F1905	Device Revision or Replacement	A070102	High Capture Threshold	28/06/2023	28/06/2023
								A0705	Battery Problem	28/06/2023	28/06/2023
								A070504	Premature Discharge of Battery	28/06/2023	28/06/2023
					CT2750	Implantable cardiac electrical stimulators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070102	High Capture Threshold	28/06/2023	28/06/2023
									A0705	Battery Problem	28/06/2023	28/06/2023
									A070504	Premature Discharge of Battery	28/06/2023	28/06/2023
							F08	Hospitalization or Prolonged Hospitalization	A070102	High Capture Threshold	28/06/2023	28/06/2023
									A0705	Battery Problem	28/06/2023	28/06/2023
									A070504	Premature Discharge of Battery	28/06/2023	28/06/2023
							F1905	Device Revision or Replacement	A070102	High Capture Threshold	28/06/2023	28/06/2023
									A0705	Battery Problem	28/06/2023	28/06/2023
									A070504	Premature Discharge of Battery	28/06/2023	28/06/2023
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A070102	High Capture Threshold	28/06/2023	28/06/2023
									A0705	Battery Problem	28/06/2023	28/06/2023
									A070504	Premature Discharge of Battery	28/06/2023	28/06/2023
							F08	Hospitalization or Prolonged Hospitalization	A070102	High Capture Threshold	28/06/2023	28/06/2023
									A0705	Battery Problem	28/06/2023	28/06/2023
									A070504	Premature Discharge of Battery	28/06/2023	28/06/2023
							F1905	Device Revision or Replacement	A070102	High Capture Threshold	28/06/2023	28/06/2023
									A0705	Battery Problem	28/06/2023	28/06/2023
									A070504	Premature Discharge of Battery	28/06/2023	28/06/2023
					CT2985	Active	E2403	No Clinical Signs, Symptoms or Conditions	A070102	High Capture Threshold	28/06/2023	28/06/2023
									A0705	Battery Problem	28/06/2023	28/06/2023
									A070504	Premature Discharge of Battery	28/06/2023	28/06/2023
							F08	Hospitalization or Prolonged Hospitalization	A070102	High Capture Threshold	28/06/2023	28/06/2023
									A0705	Battery Problem	28/06/2023	28/06/2023
									A070504	Premature Discharge of Battery	28/06/2023	28/06/2023
							F1905	Device Revision or Replacement	A070102	High Capture Threshold	28/06/2023	28/06/2023
									A0705	Battery Problem	28/06/2023	28/06/2023
									A070504	Premature Discharge of Battery	28/06/2023	28/06/2023
					CT3001	Implantable metal	E2403	No Clinical Signs, Symptoms or Conditions	A070102	High Capture Threshold	28/06/2023	28/06/2023
									A0705	Battery Problem	28/06/2023	28/06/2023
									A070504	Premature Discharge of Battery	28/06/2023	28/06/2023
							F08	Hospitalization or Prolonged Hospitalization	A070102	High Capture Threshold	28/06/2023	28/06/2023
									A0705	Battery Problem	28/06/2023	28/06/2023
									A070504	Premature Discharge of Battery	28/06/2023	28/06/2023
							F1905	Device Revision or Replacement	A070102	High Capture Threshold	28/06/2023	28/06/2023
									A0705	Battery Problem	28/06/2023	28/06/2023
									A070504	Premature Discharge of Battery	28/06/2023	28/06/2023
					CT3002	Implantable synthetic polymer	E2403	No Clinical Signs, Symptoms or Conditions	A070102	High Capture Threshold	28/06/2023	28/06/2023
									A0705	Battery Problem	28/06/2023	28/06/2023
									A070504	Premature Discharge of Battery	28/06/2023	28/06/2023
							F08	Hospitalization or Prolonged Hospitalization	A070102	High Capture Threshold	28/06/2023	28/06/2023
									A0705	Battery Problem	28/06/2023	28/06/2023
									A070504	Premature Discharge of Battery	28/06/2023	28/06/2023
							F1905	Device Revision or Replacement	A070102	High Capture Threshold	28/06/2023	28/06/2023
									A0705	Battery Problem	28/06/2023	28/06/2023
									A070504	Premature Discharge of Battery	28/06/2023	28/06/2023
					CT3006	Implantable inorganic material	E2403	No Clinical Signs, Symptoms or Conditions	A070102	High Capture Threshold	28/06/2023	28/06/2023
									A0705	Battery Problem	28/06/2023	28/06/2023
									A070504	Premature Discharge of Battery	28/06/2023	28/06/2023
							F08	Hospitalization or Prolonged Hospitalization	A070102	High Capture Threshold	28/06/2023	28/06/2023
									A0705	Battery Problem	28/06/2023	28/06/2023
									A070504	Premature Discharge of Battery	28/06/2023	28/06/2023
							F1905	Device Revision or Replacement	A070102	High Capture Threshold	28/06/2023	28/06/2023
									A0705	Battery Problem	28/06/2023	28/06/2023
									A070504	Premature Discharge of Battery	28/06/2023	28/06/2023
					CT314	Electrically powered	E2403	No Clinical Signs, Symptoms or Conditions	A070102	High Capture Threshold	28/06/2023	28/06/2023
									A0705	Battery Problem	28/06/2023	28/06/2023
									A070504	Premature Discharge of Battery	28/06/2023	28/06/2023
							F08	Hospitalization or Prolonged Hospitalization	A070102	High Capture Threshold	28/06/2023	28/06/2023
									A0705	Battery Problem	28/06/2023	28/06/2023
									A070504	Premature Discharge of Battery	28/06/2023	28/06/2023
							F1905	Device Revision or Replacement	A070102	High Capture Threshold	28/06/2023	28/06/2023
									A0705	Battery Problem	28/06/2023	28/06/2023
									A070504	Premature Discharge of Battery	28/06/2023	28/06/2023
					CT321	Long-term surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A070102	High Capture Threshold	28/06/2023	28/06/2023
									A0705	Battery Problem	28/06/2023	28/06/2023
									A070504	Premature Discharge of Battery	28/06/2023	28/06/2023
							F08	Hospitalization or Prolonged Hospitalization	A070102	High Capture Threshold	28/06/2023	28/06/2023
									A0705	Battery Problem	28/06/2023	28/06/2023
									A070504	Premature Discharge of Battery	28/06/2023	28/06/2023
							F1905	Device Revision or Replacement	A070102	High Capture Threshold	28/06/2023	28/06/2023
									A0705	Battery Problem	28/06/2023	28/06/2023
									A070504	Premature Discharge of Battery	28/06/2023	28/06/2023
					CT449	Generators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070102	High Capture Threshold	28/06/2023	28/06/2023
									A0705	Battery Problem	28/06/2023	28/06/2023
									A070504	Premature Discharge of Battery	28/06/2023	28/06/2023
							F08	Hospitalization or Prolonged Hospitalization	A070102	High Capture Threshold	28/06/2023	28/06/2023
									A0705	Battery Problem	28/06/2023	28/06/2023
									A070504	Premature Discharge of Battery	28/06/2023	28/06/2023
							F1905	Device Revision or Replacement	A070102	High Capture Threshold	28/06/2023	28/06/2023
									A0705	Battery Problem	28/06/2023	28/06/2023
									A070504	Premature Discharge of Battery	28/06/2023	28/06/2023
					CT981	Single use	E2403	No Clinical Signs, Symptoms or Conditions	A070102	High Capture Threshold	28/06/2023	28/06/2023
									A0705	Battery Problem	28/06/2023	28/06/2023
									A070504	Premature Discharge of Battery	28/06/2023	28/06/2023

							F08	Hospitalization or Prolonged Hospitalization	A070102	High Capture Threshold	28/06/2023	28/06/2023
									A0705	Battery Problem	28/06/2023	28/06/2023
									A070504	Premature Discharge of Battery	28/06/2023	28/06/2023
							F1905	Device Revision or Replacement	A070102	High Capture Threshold	28/06/2023	28/06/2023
									A0705	Battery Problem	28/06/2023	28/06/2023
									A070504	Premature Discharge of Battery	28/06/2023	28/06/2023
					CT983	Surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A070102	High Capture Threshold	28/06/2023	28/06/2023
									A0705	Battery Problem	28/06/2023	28/06/2023
									A070504	Premature Discharge of Battery	28/06/2023	28/06/2023
							F08	Hospitalization or Prolonged Hospitalization	A070102	High Capture Threshold	28/06/2023	28/06/2023
									A0705	Battery Problem	28/06/2023	28/06/2023
									A070504	Premature Discharge of Battery	28/06/2023	28/06/2023
							F1905	Device Revision or Replacement	A070102	High Capture Threshold	28/06/2023	28/06/2023
									A0705	Battery Problem	28/06/2023	28/06/2023
									A070504	Premature Discharge of Battery	28/06/2023	28/06/2023
14147	2023	All other reportable incidents	47265	Dual-chamber implantable pacemaker, rate-responsive	CT141	Electrodes/Leads and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070909	Oversensing		
									A0712	Pacing Problem		
									A090801	Signal Artifact/Noise		
							F1001	Absence of Treatment	A070909	Oversensing		
									A0712	Pacing Problem		
									A090801	Signal Artifact/Noise		
							F2203	Imaging Required	A070909	Oversensing		
									A0712	Pacing Problem		
									A090801	Signal Artifact/Noise		
					CT145	Cardiovascular devices	E2403	No Clinical Signs, Symptoms or Conditions	A070909	Oversensing		
									A0712	Pacing Problem		
									A090801	Signal Artifact/Noise		
							F1001	Absence of Treatment	A070909	Oversensing		
									A0712	Pacing Problem		
									A090801	Signal Artifact/Noise		
							F2203	Imaging Required	A070909	Oversensing		
									A0712	Pacing Problem		
									A090801	Signal Artifact/Noise		
					CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070909	Oversensing		
									A0712	Pacing Problem		
									A090801	Signal Artifact/Noise		
							F1001	Absence of Treatment	A070909	Oversensing		
									A0712	Pacing Problem		
									A090801	Signal Artifact/Noise		
							F2203	Imaging Required	A070909	Oversensing		
									A0712	Pacing Problem		
									A090801	Signal Artifact/Noise		
					CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A070909	Oversensing		
									A0712	Pacing Problem		
									A090801	Signal Artifact/Noise		
							F1001	Absence of Treatment	A070909	Oversensing		
									A0712	Pacing Problem		
									A090801	Signal Artifact/Noise		
							F2203	Imaging Required	A070909	Oversensing		
									A0712	Pacing Problem		
									A090801	Signal Artifact/Noise		
					CT233	Surgical	E2403	No Clinical Signs, Symptoms or Conditions	A070909	Oversensing		
									A0712	Pacing Problem		
									A090801	Signal Artifact/Noise		
							F1001	Absence of Treatment	A070909	Oversensing		
									A0712	Pacing Problem		
									A090801	Signal Artifact/Noise		
							F2203	Imaging Required	A070909	Oversensing		
									A0712	Pacing Problem		
									A090801	Signal Artifact/Noise		
					CT2406	Implantable	E2403	No Clinical Signs, Symptoms or Conditions	A070909	Oversensing		
									A0712	Pacing Problem		
									A090801	Signal Artifact/Noise		
							F1001	Absence of Treatment	A070909	Oversensing		
									A0712	Pacing Problem		
									A090801	Signal Artifact/Noise		
							F2203	Imaging Required	A070909	Oversensing		
									A0712	Pacing Problem		
									A090801	Signal Artifact/Noise		
					CT2407	Active implantable	E2403	No Clinical Signs, Symptoms or Conditions	A070909	Oversensing		
									A0712	Pacing Problem		
									A090801	Signal Artifact/Noise		
							F1001	Absence of Treatment	A070909	Oversensing		
									A0712	Pacing Problem		
									A090801	Signal Artifact/Noise		
							F2203	Imaging Required	A070909	Oversensing		
									A0712	Pacing Problem		
									A090801	Signal Artifact/Noise		
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A070909	Oversensing		
									A0712	Pacing Problem		
									A090801	Signal Artifact/Noise		
							F1001	Absence of Treatment	A070909	Oversensing		
									A0712	Pacing Problem		
									A090801	Signal Artifact/Noise		
							F2203	Imaging Required	A070909	Oversensing		
									A0712	Pacing Problem		
									A090801	Signal Artifact/Noise		
					CT2727	Active implant devices/ancillaries	E2403	No Clinical Signs, Symptoms or Conditions	A070909	Oversensing		

									A0712	Pacing Problem		
									A090801	Signal Artifact/Noise		
							F1001	Absence of Treatment	A070909	Oversensing		
									A0712	Pacing Problem		
									A090801	Signal Artifact/Noise		
							F2203	Imaging Required	A070909	Oversensing		
									A0712	Pacing Problem		
									A090801	Signal Artifact/Noise		
					CT2750	Implantable cardiac electrical stimulators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070909	Oversensing		
									A0712	Pacing Problem		
									A090801	Signal Artifact/Noise		
							F1001	Absence of Treatment	A070909	Oversensing		
									A0712	Pacing Problem		
									A090801	Signal Artifact/Noise		
							F2203	Imaging Required	A070909	Oversensing		
									A0712	Pacing Problem		
									A090801	Signal Artifact/Noise		
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A070909	Oversensing		
									A0712	Pacing Problem		
									A090801	Signal Artifact/Noise		
							F1001	Absence of Treatment	A070909	Oversensing		
									A0712	Pacing Problem		
									A090801	Signal Artifact/Noise		
							F2203	Imaging Required	A070909	Oversensing		
									A0712	Pacing Problem		
									A090801	Signal Artifact/Noise		
					CT2985	Active	E2403	No Clinical Signs, Symptoms or Conditions	A070909	Oversensing		
									A0712	Pacing Problem		
									A090801	Signal Artifact/Noise		
							F1001	Absence of Treatment	A070909	Oversensing		
									A0712	Pacing Problem		
									A090801	Signal Artifact/Noise		
							F2203	Imaging Required	A070909	Oversensing		
									A0712	Pacing Problem		
									A090801	Signal Artifact/Noise		
					CT3001	Implantable metal	E2403	No Clinical Signs, Symptoms or Conditions	A070909	Oversensing		
									A0712	Pacing Problem		
									A090801	Signal Artifact/Noise		
							F1001	Absence of Treatment	A070909	Oversensing		
									A0712	Pacing Problem		
									A090801	Signal Artifact/Noise		
							F2203	Imaging Required	A070909	Oversensing		
									A0712	Pacing Problem		
									A090801	Signal Artifact/Noise		
					CT3002	Implantable synthetic polymer	E2403	No Clinical Signs, Symptoms or Conditions	A070909	Oversensing		
									A0712	Pacing Problem		
									A090801	Signal Artifact/Noise		
							F1001	Absence of Treatment	A070909	Oversensing		
									A0712	Pacing Problem		
									A090801	Signal Artifact/Noise		
							F2203	Imaging Required	A070909	Oversensing		
									A0712	Pacing Problem		
									A090801	Signal Artifact/Noise		
					CT3006	Implantable inorganic material	E2403	No Clinical Signs, Symptoms or Conditions	A070909	Oversensing		
									A0712	Pacing Problem		
									A090801	Signal Artifact/Noise		
							F1001	Absence of Treatment	A070909	Oversensing		
									A0712	Pacing Problem		
									A090801	Signal Artifact/Noise		
							F2203	Imaging Required	A070909	Oversensing		
									A0712	Pacing Problem		
									A090801	Signal Artifact/Noise		
					CT314	Electrically powered	E2403	No Clinical Signs, Symptoms or Conditions	A070909	Oversensing		
									A0712	Pacing Problem		
									A090801	Signal Artifact/Noise		
							F1001	Absence of Treatment	A070909	Oversensing		
									A0712	Pacing Problem		
									A090801	Signal Artifact/Noise		
							F2203	Imaging Required	A070909	Oversensing		
									A0712	Pacing Problem		
									A090801	Signal Artifact/Noise		
					CT321	Long-term surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A070909	Oversensing		
									A0712	Pacing Problem		
									A090801	Signal Artifact/Noise		
							F1001	Absence of Treatment	A070909	Oversensing		
									A0712	Pacing Problem		
									A090801	Signal Artifact/Noise		
							F2203	Imaging Required	A070909	Oversensing		
									A0712	Pacing Problem		
									A090801	Signal Artifact/Noise		
					CT449	Generators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070909	Oversensing		
									A0712	Pacing Problem		
									A090801	Signal Artifact/Noise		
							F1001	Absence of Treatment	A070909	Oversensing		
									A0712	Pacing Problem		
									A090801	Signal Artifact/Noise		
							F2203	Imaging Required	A070909	Oversensing		
									A0712	Pacing Problem		

								A090801	Signal Artifact/Noise			
				CT981	Single use	E2403	No Clinical Signs, Symptoms or Conditions	A070909	Oversensing			
								A0712	Pacing Problem			
								A090801	Signal Artifact/Noise			
						F1001	Absence of Treatment	A070909	Oversensing			
								A0712	Pacing Problem			
								A090801	Signal Artifact/Noise			
						F2203	Imaging Required	A070909	Oversensing			
								A0712	Pacing Problem			
								A090801	Signal Artifact/Noise			
				CT983	Surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A070909	Oversensing			
								A0712	Pacing Problem			
								A090801	Signal Artifact/Noise			
						F1001	Absence of Treatment	A070909	Oversensing			
								A0712	Pacing Problem			
								A090801	Signal Artifact/Noise			
						F2203	Imaging Required	A070909	Oversensing			
								A0712	Pacing Problem			
								A090801	Signal Artifact/Noise			
14148	2023		47265	Dual-chamber implantable pacemaker, rate-responsive	CT141	Electrodes/Leads and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	20/09/2023	20/09/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	20/09/2023	20/09/2023
					CT145	Cardiovascular devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	20/09/2023	20/09/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	20/09/2023	20/09/2023
					CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	20/09/2023	20/09/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	20/09/2023	20/09/2023
					CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	20/09/2023	20/09/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	20/09/2023	20/09/2023
					CT233	Surgical	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	20/09/2023	20/09/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	20/09/2023	20/09/2023
					CT2406	Implantable	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	20/09/2023	20/09/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	20/09/2023	20/09/2023
					CT2407	Active implantable	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	20/09/2023	20/09/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	20/09/2023	20/09/2023
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	20/09/2023	20/09/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	20/09/2023	20/09/2023
					CT2727	Active implant devices/ancillaries	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	20/09/2023	20/09/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	20/09/2023	20/09/2023
					CT2750	Implantable cardiac electrical stimulators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	20/09/2023	20/09/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	20/09/2023	20/09/2023
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	20/09/2023	20/09/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	20/09/2023	20/09/2023
					CT2985	Active	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	20/09/2023	20/09/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	20/09/2023	20/09/2023
					CT3001	Implantable metal	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	20/09/2023	20/09/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	20/09/2023	20/09/2023
					CT3002	Implantable synthetic polymer	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	20/09/2023	20/09/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	20/09/2023	20/09/2023
					CT3006	Implantable inorganic material	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	20/09/2023	20/09/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	20/09/2023	20/09/2023
					CT314	Electrically powered	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	20/09/2023	20/09/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	20/09/2023	20/09/2023
					CT321	Long-term surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	20/09/2023	20/09/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	20/09/2023	20/09/2023
					CT449	Generators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	20/09/2023	20/09/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	20/09/2023	20/09/2023
					CT981	Single use	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	20/09/2023	20/09/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	20/09/2023	20/09/2023
					CT983	Surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	20/09/2023	20/09/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	20/09/2023	20/09/2023
14149	2023		47265	Dual-chamber implantable pacemaker, rate-responsive	CT141	Electrodes/Leads and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	02/10/2023	02/10/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	02/10/2023	02/10/2023
					CT145	Cardiovascular devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	02/10/2023	02/10/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	02/10/2023	02/10/2023
					CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	02/10/2023	02/10/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	02/10/2023	02/10/2023
					CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	02/10/2023	02/10/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	02/10/2023	02/10/2023
					CT233	Surgical	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	02/10/2023	02/10/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	02/10/2023	02/10/2023
					CT2406	Implantable	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	02/10/2023	02/10/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	02/10/2023	02/10/2023
					CT2407	Active implantable	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	02/10/2023	02/10/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	02/10/2023	02/10/2023
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	02/10/2023	02/10/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	02/10/2023	02/10/2023
					CT2727	Active implant devices/ancillaries	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	02/10/2023	02/10/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	02/10/2023	02/10/2023
					CT2750	Implantable cardiac electrical stimulators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	02/10/2023	02/10/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	02/10/2023	02/10/2023
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	02/10/2023	02/10/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	02/10/2023	02/10/2023
					CT2985	Active	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	02/10/2023	02/10/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	02/10/2023	02/10/2023
					CT3001	Implantable metal	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	02/10/2023	02/10/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	02/10/2023	02/10/2023
					CT3002	Implantable synthetic polymer	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	02/10/2023	02/10/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	02/10/2023	02/10/2023
					CT3006	Implantable inorganic material	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	02/10/2023	02/10/2023

							F1901	Additional Surgery	A24	No Apparent Device Problem	02/10/2023	02/10/2023
					CT314	Electrically powered	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	02/10/2023	02/10/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	02/10/2023	02/10/2023
					CT321	Long-term surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	02/10/2023	02/10/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	02/10/2023	02/10/2023
					CT449	Generators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	02/10/2023	02/10/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	02/10/2023	02/10/2023
					CT981	Single use	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	02/10/2023	02/10/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	02/10/2023	02/10/2023
					CT983	Surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	02/10/2023	02/10/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	02/10/2023	02/10/2023
14150	2023		47265	Dual-chamber implantable pacemaker, rate-responsive	CT141	Electrodes/Leads and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	02/10/2023	02/10/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	02/10/2023	02/10/2023
					CT145	Cardiovascular devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	02/10/2023	02/10/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	02/10/2023	02/10/2023
					CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	02/10/2023	02/10/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	02/10/2023	02/10/2023
					CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	02/10/2023	02/10/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	02/10/2023	02/10/2023
					CT233	Surgical	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	02/10/2023	02/10/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	02/10/2023	02/10/2023
					CT2406	Implantable	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	02/10/2023	02/10/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	02/10/2023	02/10/2023
					CT2407	Active implantable	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	02/10/2023	02/10/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	02/10/2023	02/10/2023
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	02/10/2023	02/10/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	02/10/2023	02/10/2023
					CT2727	Active implant devices/ancillaries	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	02/10/2023	02/10/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	02/10/2023	02/10/2023
					CT2750	Implantable cardiac electrical stimulators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	02/10/2023	02/10/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	02/10/2023	02/10/2023
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	02/10/2023	02/10/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	02/10/2023	02/10/2023
					CT2985	Active	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	02/10/2023	02/10/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	02/10/2023	02/10/2023
					CT3001	Implantable metal	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	02/10/2023	02/10/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	02/10/2023	02/10/2023
					CT3002	Implantable synthetic polymer	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	02/10/2023	02/10/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	02/10/2023	02/10/2023
					CT3006	Implantable inorganic material	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	02/10/2023	02/10/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	02/10/2023	02/10/2023
					CT314	Electrically powered	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	02/10/2023	02/10/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	02/10/2023	02/10/2023
					CT321	Long-term surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	02/10/2023	02/10/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	02/10/2023	02/10/2023
					CT449	Generators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	02/10/2023	02/10/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	02/10/2023	02/10/2023
					CT981	Single use	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	02/10/2023	02/10/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	02/10/2023	02/10/2023
					CT983	Surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	02/10/2023	02/10/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	02/10/2023	02/10/2023
14151	2023		47265	Dual-chamber implantable pacemaker, rate-responsive	CT141	Electrodes/Leads and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	03/10/2023	03/10/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	03/10/2023	03/10/2023
					CT145	Cardiovascular devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	03/10/2023	03/10/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	03/10/2023	03/10/2023
					CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	03/10/2023	03/10/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	03/10/2023	03/10/2023
					CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	03/10/2023	03/10/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	03/10/2023	03/10/2023
					CT233	Surgical	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	03/10/2023	03/10/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	03/10/2023	03/10/2023
					CT2406	Implantable	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	03/10/2023	03/10/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	03/10/2023	03/10/2023
					CT2407	Active implantable	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	03/10/2023	03/10/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	03/10/2023	03/10/2023
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	03/10/2023	03/10/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	03/10/2023	03/10/2023
					CT2727	Active implant devices/ancillaries	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	03/10/2023	03/10/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	03/10/2023	03/10/2023
					CT2750	Implantable cardiac electrical stimulators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	03/10/2023	03/10/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	03/10/2023	03/10/2023
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	03/10/2023	03/10/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	03/10/2023	03/10/2023
					CT2985	Active	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	03/10/2023	03/10/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	03/10/2023	03/10/2023
					CT3001	Implantable metal	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	03/10/2023	03/10/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	03/10/2023	03/10/2023
					CT3002	Implantable synthetic polymer	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	03/10/2023	03/10/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	03/10/2023	03/10/2023
					CT3006	Implantable inorganic material	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	03/10/2023	03/10/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	03/10/2023	03/10/2023
					CT314	Electrically powered	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	03/10/2023	03/10/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	03/10/2023	03/10/2023
					CT321	Long-term surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	03/10/2023	03/10/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	03/10/2023	03/10/2023
					CT449	Generators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	03/10/2023	03/10/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	03/10/2023	03/10/2023
					CT981	Single use	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	03/10/2023	03/10/2023

						F1901	Additional Surgery	A24	No Apparent Device Problem	03/10/2023	03/10/2023
				CT983	Surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	03/10/2023	03/10/2023
						F1901	Additional Surgery	A24	No Apparent Device Problem	03/10/2023	03/10/2023
14152	2023		47265	Dual-chamber implantable pacemaker, rate-responsive	CT141	Electrodes/Leads and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	03/10/2023	03/10/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	03/10/2023	03/10/2023
					CT145	Cardiovascular devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	03/10/2023	03/10/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	03/10/2023	03/10/2023
					CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	03/10/2023	03/10/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	03/10/2023	03/10/2023
					CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	03/10/2023	03/10/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	03/10/2023	03/10/2023
					CT233	Surgical	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	03/10/2023	03/10/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	03/10/2023	03/10/2023
					CT2406	Implantable	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	03/10/2023	03/10/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	03/10/2023	03/10/2023
					CT2407	Active implantable	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	03/10/2023	03/10/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	03/10/2023	03/10/2023
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	03/10/2023	03/10/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	03/10/2023	03/10/2023
					CT2727	Active implant devices/ancillaries	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	03/10/2023	03/10/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	03/10/2023	03/10/2023
					CT2750	Implantable cardiac electrical stimulators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	03/10/2023	03/10/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	03/10/2023	03/10/2023
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	03/10/2023	03/10/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	03/10/2023	03/10/2023
					CT2985	Active	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	03/10/2023	03/10/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	03/10/2023	03/10/2023
					CT3001	Implantable metal	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	03/10/2023	03/10/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	03/10/2023	03/10/2023
					CT3002	Implantable synthetic polymer	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	03/10/2023	03/10/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	03/10/2023	03/10/2023
					CT3006	Implantable inorganic material	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	03/10/2023	03/10/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	03/10/2023	03/10/2023
					CT314	Electrically powered	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	03/10/2023	03/10/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	03/10/2023	03/10/2023
					CT321	Long-term surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	03/10/2023	03/10/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	03/10/2023	03/10/2023
					CT449	Generators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	03/10/2023	03/10/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	03/10/2023	03/10/2023
					CT981	Single use	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	03/10/2023	03/10/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	03/10/2023	03/10/2023
					CT983	Surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	03/10/2023	03/10/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	03/10/2023	03/10/2023
14153	2023		47265	Dual-chamber implantable pacemaker, rate-responsive	CT141	Electrodes/Leads and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	03/10/2023	03/10/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	03/10/2023	03/10/2023
					CT145	Cardiovascular devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	03/10/2023	03/10/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	03/10/2023	03/10/2023
					CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	03/10/2023	03/10/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	03/10/2023	03/10/2023
					CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	03/10/2023	03/10/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	03/10/2023	03/10/2023
					CT233	Surgical	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	03/10/2023	03/10/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	03/10/2023	03/10/2023
					CT2406	Implantable	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	03/10/2023	03/10/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	03/10/2023	03/10/2023
					CT2407	Active implantable	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	03/10/2023	03/10/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	03/10/2023	03/10/2023
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	03/10/2023	03/10/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	03/10/2023	03/10/2023
					CT2727	Active implant devices/ancillaries	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	03/10/2023	03/10/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	03/10/2023	03/10/2023
					CT2750	Implantable cardiac electrical stimulators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	03/10/2023	03/10/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	03/10/2023	03/10/2023
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	03/10/2023	03/10/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	03/10/2023	03/10/2023
					CT2985	Active	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	03/10/2023	03/10/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	03/10/2023	03/10/2023
					CT3001	Implantable metal	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	03/10/2023	03/10/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	03/10/2023	03/10/2023
					CT3002	Implantable synthetic polymer	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	03/10/2023	03/10/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	03/10/2023	03/10/2023
					CT3006	Implantable inorganic material	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	03/10/2023	03/10/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	03/10/2023	03/10/2023
					CT314	Electrically powered	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	03/10/2023	03/10/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	03/10/2023	03/10/2023
					CT321	Long-term surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	03/10/2023	03/10/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	03/10/2023	03/10/2023
					CT449	Generators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	03/10/2023	03/10/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	03/10/2023	03/10/2023
					CT981	Single use	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	03/10/2023	03/10/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	03/10/2023	03/10/2023
					CT983	Surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	03/10/2023	03/10/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	03/10/2023	03/10/2023
14154	2023		47265	Dual-chamber implantable pacemaker, rate-responsive	CT141	Electrodes/Leads and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	03/10/2023	03/10/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	03/10/2023	03/10/2023
					CT145	Cardiovascular devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	03/10/2023	03/10/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	03/10/2023	03/10/2023
					CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	03/10/2023	03/10/2023

							F1901	Additional Surgery	A24	No Apparent Device Problem	03/10/2023	03/10/2023
					CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	03/10/2023	03/10/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	03/10/2023	03/10/2023
					CT233	Surgical	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	03/10/2023	03/10/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	03/10/2023	03/10/2023
					CT2406	Implantable	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	03/10/2023	03/10/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	03/10/2023	03/10/2023
					CT2407	Active implantable	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	03/10/2023	03/10/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	03/10/2023	03/10/2023
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	03/10/2023	03/10/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	03/10/2023	03/10/2023
					CT2727	Active implant devices/ancillaries	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	03/10/2023	03/10/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	03/10/2023	03/10/2023
					CT2750	Implantable cardiac electrical stimulators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	03/10/2023	03/10/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	03/10/2023	03/10/2023
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	03/10/2023	03/10/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	03/10/2023	03/10/2023
					CT2985	Active	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	03/10/2023	03/10/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	03/10/2023	03/10/2023
					CT3001	Implantable metal	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	03/10/2023	03/10/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	03/10/2023	03/10/2023
					CT3002	Implantable synthetic polymer	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	03/10/2023	03/10/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	03/10/2023	03/10/2023
					CT3006	Implantable inorganic material	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	03/10/2023	03/10/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	03/10/2023	03/10/2023
					CT314	Electrically powered	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	03/10/2023	03/10/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	03/10/2023	03/10/2023
					CT321	Long-term surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	03/10/2023	03/10/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	03/10/2023	03/10/2023
					CT449	Generators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	03/10/2023	03/10/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	03/10/2023	03/10/2023
					CT981	Single use	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	03/10/2023	03/10/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	03/10/2023	03/10/2023
					CT983	Surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	03/10/2023	03/10/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	03/10/2023	03/10/2023
14155	2023	All other reportable incidents	17882	Physiologic-monitoring defibrillation system	CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070603	Failure to Charge		
									A071204	No Pacing		
							F27	Problem Identified During Non-Clinical Procedure	A070603	Failure to Charge		
									A071204	No Pacing		
14156	2023		47265	Dual-chamber implantable pacemaker, rate-responsive	CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	02/10/2023	02/10/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	02/10/2023	02/10/2023
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	02/10/2023	02/10/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	02/10/2023	02/10/2023
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	02/10/2023	02/10/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	02/10/2023	02/10/2023
14157	2023		47265	Dual-chamber implantable pacemaker, rate-responsive	CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	02/10/2023	02/10/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	02/10/2023	02/10/2023
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	02/10/2023	02/10/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	02/10/2023	02/10/2023
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	02/10/2023	02/10/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	02/10/2023	02/10/2023
14158	2023	All other reportable incidents	36241	Pacemaker/defibrillator lead	CT1684	Cardiac pacemakers and associated devices	E060101	Asystole	A070909	Oversensing		
									A0712	Pacing Problem		
									A072201	High Impedance		
							F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A070909	Oversensing		
									A0712	Pacing Problem		
									A072201	High Impedance		
							F1001	Absence of Treatment	A070909	Oversensing		
									A0712	Pacing Problem		
									A072201	High Impedance		
							F15	Recognized Device or Procedural Complication	A070909	Oversensing		
									A0712	Pacing Problem		
									A072201	High Impedance		
14159	2023	All other reportable incidents	47263	Cardiac resynchronization therapy implantable pacemaker	CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A0908	Incorrect, Inadequate or Imprecise Result or Readings	31/07/2023	31/07/2023
							F08	Hospitalization or Prolonged Hospitalization	A0908	Incorrect, Inadequate or Imprecise Result or Readings	31/07/2023	31/07/2023
							F1905	Device Revision or Replacement	A0908	Incorrect, Inadequate or Imprecise Result or Readings	31/07/2023	31/07/2023
					CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A0908	Incorrect, Inadequate or Imprecise Result or Readings	31/07/2023	31/07/2023
							F08	Hospitalization or Prolonged Hospitalization	A0908	Incorrect, Inadequate or Imprecise Result or Readings	31/07/2023	31/07/2023
							F1905	Device Revision or Replacement	A0908	Incorrect, Inadequate or Imprecise Result or Readings	31/07/2023	31/07/2023
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A0908	Incorrect, Inadequate or Imprecise Result or Readings	31/07/2023	31/07/2023
							F08	Hospitalization or Prolonged Hospitalization	A0908	Incorrect, Inadequate or Imprecise Result or Readings	31/07/2023	31/07/2023
							F1905	Device Revision or Replacement	A0908	Incorrect, Inadequate or Imprecise Result or Readings	31/07/2023	31/07/2023
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A0908	Incorrect, Inadequate or Imprecise Result or Readings	31/07/2023	31/07/2023
							F08	Hospitalization or Prolonged Hospitalization	A0908	Incorrect, Inadequate or Imprecise Result or Readings	31/07/2023	31/07/2023
							F1905	Device Revision or Replacement	A0908	Incorrect, Inadequate or Imprecise Result or Readings	31/07/2023	31/07/2023
14160	2023		47265	Dual-chamber implantable pacemaker, rate-responsive	CT141	Electrodes/Leads and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	04/06/2023	04/06/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	04/06/2023	04/06/2023
					CT145	Cardiovascular devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	04/06/2023	04/06/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	04/06/2023	04/06/2023

					CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	04/06/2023	04/06/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	04/06/2023	04/06/2023
					CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	04/06/2023	04/06/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	04/06/2023	04/06/2023
					CT233	Surgical	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	04/06/2023	04/06/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	04/06/2023	04/06/2023
					CT2406	Implantable	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	04/06/2023	04/06/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	04/06/2023	04/06/2023
					CT2407	Active implantable	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	04/06/2023	04/06/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	04/06/2023	04/06/2023
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	04/06/2023	04/06/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	04/06/2023	04/06/2023
					CT2727	Active implant devices/ancillaries	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	04/06/2023	04/06/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	04/06/2023	04/06/2023
					CT2750	Implantable cardiac electrical stimulators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	04/06/2023	04/06/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	04/06/2023	04/06/2023
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	04/06/2023	04/06/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	04/06/2023	04/06/2023
					CT2985	Active	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	04/06/2023	04/06/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	04/06/2023	04/06/2023
					CT3001	Implantable metal	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	04/06/2023	04/06/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	04/06/2023	04/06/2023
					CT3002	Implantable synthetic polymer	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	04/06/2023	04/06/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	04/06/2023	04/06/2023
					CT3006	Implantable inorganic material	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	04/06/2023	04/06/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	04/06/2023	04/06/2023
					CT314	Electrically powered	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	04/06/2023	04/06/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	04/06/2023	04/06/2023
					CT321	Long-term surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	04/06/2023	04/06/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	04/06/2023	04/06/2023
					CT449	Generators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	04/06/2023	04/06/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	04/06/2023	04/06/2023
					CT981	Single use	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	04/06/2023	04/06/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	04/06/2023	04/06/2023
					CT983	Surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	04/06/2023	04/06/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	04/06/2023	04/06/2023
14161	2023	All other reportable incidents	60789	Intracardiac pacemaker	CT130	Catheters and associated devices	E060106	Heart Block	A070102	High Capture Threshold		
							E2403	No Clinical Signs, Symptoms or Conditions	A070102	High Capture Threshold		
							F12	Serious Injury/ Illness/ Impairment	A070102	High Capture Threshold		
							F23	Unexpected Medical Intervention	A070102	High Capture Threshold		
					CT145	Cardiovascular devices	E060106	Heart Block	A070102	High Capture Threshold		
							E2403	No Clinical Signs, Symptoms or Conditions	A070102	High Capture Threshold		
							F12	Serious Injury/ Illness/ Impairment	A070102	High Capture Threshold		
							F23	Unexpected Medical Intervention	A070102	High Capture Threshold		
					CT1684	Cardiac pacemakers and associated devices	E060106	Heart Block	A070102	High Capture Threshold		
							E2403	No Clinical Signs, Symptoms or Conditions	A070102	High Capture Threshold		
							F12	Serious Injury/ Illness/ Impairment	A070102	High Capture Threshold		
							F23	Unexpected Medical Intervention	A070102	High Capture Threshold		
					CT1685	Cardiac pacemakers	E060106	Heart Block	A070102	High Capture Threshold		
							E2403	No Clinical Signs, Symptoms or Conditions	A070102	High Capture Threshold		
							F12	Serious Injury/ Illness/ Impairment	A070102	High Capture Threshold		
							F23	Unexpected Medical Intervention	A070102	High Capture Threshold		
					CT233	Surgical	E060106	Heart Block	A070102	High Capture Threshold		
							E2403	No Clinical Signs, Symptoms or Conditions	A070102	High Capture Threshold		
							F12	Serious Injury/ Illness/ Impairment	A070102	High Capture Threshold		
							F23	Unexpected Medical Intervention	A070102	High Capture Threshold		
					CT2406	Implantable	E060106	Heart Block	A070102	High Capture Threshold		
							E2403	No Clinical Signs, Symptoms or Conditions	A070102	High Capture Threshold		
							F12	Serious Injury/ Illness/ Impairment	A070102	High Capture Threshold		
							F23	Unexpected Medical Intervention	A070102	High Capture Threshold		
					CT2407	Active implantable	E060106	Heart Block	A070102	High Capture Threshold		
							E2403	No Clinical Signs, Symptoms or Conditions	A070102	High Capture Threshold		
							F12	Serious Injury/ Illness/ Impairment	A070102	High Capture Threshold		
							F23	Unexpected Medical Intervention	A070102	High Capture Threshold		
					CT257	Implantable cardiac pacemakers	E060106	Heart Block	A070102	High Capture Threshold		
							E2403	No Clinical Signs, Symptoms or Conditions	A070102	High Capture Threshold		
							F12	Serious Injury/ Illness/ Impairment	A070102	High Capture Threshold		
							F23	Unexpected Medical Intervention	A070102	High Capture Threshold		
					CT2727	Active implant devices/ancillaries	E060106	Heart Block	A070102	High Capture Threshold		
							E2403	No Clinical Signs, Symptoms or Conditions	A070102	High Capture Threshold		
							F12	Serious Injury/ Illness/ Impairment	A070102	High Capture Threshold		
							F23	Unexpected Medical Intervention	A070102	High Capture Threshold		
					CT2750	Implantable cardiac electrical stimulators and associated devices	E060106	Heart Block	A070102	High Capture Threshold		
							E2403	No Clinical Signs, Symptoms or Conditions	A070102	High Capture Threshold		
							F12	Serious Injury/ Illness/ Impairment	A070102	High Capture Threshold		
							F23	Unexpected Medical Intervention	A070102	High Capture Threshold		
					CT2818	Catheterization/Cannulation/Fluid path devices	E060106	Heart Block	A070102	High Capture Threshold		
							E2403	No Clinical Signs, Symptoms or Conditions	A070102	High Capture Threshold		
							F12	Serious Injury/ Illness/ Impairment	A070102	High Capture Threshold		
							F23	Unexpected Medical Intervention	A070102	High Capture Threshold		
					CT2866	Implantable cardiac electrical stimulators	E060106	Heart Block	A070102	High Capture Threshold		
							E2403	No Clinical Signs, Symptoms or Conditions	A070102	High Capture Threshold		
							F12	Serious Injury/ Illness/ Impairment	A070102	High Capture Threshold		
							F23	Unexpected Medical Intervention	A070102	High Capture Threshold		
					CT2985	Active	E060106	Heart Block	A070102	High Capture Threshold		
							E2403	No Clinical Signs, Symptoms or Conditions	A070102	High Capture Threshold		
							F12	Serious Injury/ Illness/ Impairment	A070102	High Capture Threshold		
							F23	Unexpected Medical Intervention	A070102	High Capture Threshold		

					CT3001	Implantable metal	E060106	Heart Block	A070102	High Capture Threshold		
							E2403	No Clinical Signs, Symptoms or Conditions	A070102	High Capture Threshold		
							F12	Serious Injury/ Illness/ Impairment	A070102	High Capture Threshold		
							F23	Unexpected Medical Intervention	A070102	High Capture Threshold		
					CT3002	Implantable synthetic polymer	E060106	Heart Block	A070102	High Capture Threshold		
							E2403	No Clinical Signs, Symptoms or Conditions	A070102	High Capture Threshold		
							F12	Serious Injury/ Illness/ Impairment	A070102	High Capture Threshold		
							F23	Unexpected Medical Intervention	A070102	High Capture Threshold		
					CT3006	Implantable inorganic material	E060106	Heart Block	A070102	High Capture Threshold		
							E2403	No Clinical Signs, Symptoms or Conditions	A070102	High Capture Threshold		
							F12	Serious Injury/ Illness/ Impairment	A070102	High Capture Threshold		
							F23	Unexpected Medical Intervention	A070102	High Capture Threshold		
					CT314	Electrically powered	E060106	Heart Block	A070102	High Capture Threshold		
							E2403	No Clinical Signs, Symptoms or Conditions	A070102	High Capture Threshold		
							F12	Serious Injury/ Illness/ Impairment	A070102	High Capture Threshold		
							F23	Unexpected Medical Intervention	A070102	High Capture Threshold		
					CT321	Long-term surgical invasive	E060106	Heart Block	A070102	High Capture Threshold		
							E2403	No Clinical Signs, Symptoms or Conditions	A070102	High Capture Threshold		
							F12	Serious Injury/ Illness/ Impairment	A070102	High Capture Threshold		
							F23	Unexpected Medical Intervention	A070102	High Capture Threshold		
					CT981	Single use	E060106	Heart Block	A070102	High Capture Threshold		
							E2403	No Clinical Signs, Symptoms or Conditions	A070102	High Capture Threshold		
							F12	Serious Injury/ Illness/ Impairment	A070102	High Capture Threshold		
							F23	Unexpected Medical Intervention	A070102	High Capture Threshold		
					CT983	Surgical invasive	E060106	Heart Block	A070102	High Capture Threshold		
							E2403	No Clinical Signs, Symptoms or Conditions	A070102	High Capture Threshold		
							F12	Serious Injury/ Illness/ Impairment	A070102	High Capture Threshold		
							F23	Unexpected Medical Intervention	A070102	High Capture Threshold		
14162	2023		47265	Dual-chamber implantable pacemaker, rate-responsive	CT141	Electrodes/Leads and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	28/09/2023	28/09/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	28/09/2023	28/09/2023
					CT145	Cardiovascular devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	28/09/2023	28/09/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	28/09/2023	28/09/2023
					CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	28/09/2023	28/09/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	28/09/2023	28/09/2023
					CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	28/09/2023	28/09/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	28/09/2023	28/09/2023
					CT233	Surgical	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	28/09/2023	28/09/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	28/09/2023	28/09/2023
					CT2406	Implantable	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	28/09/2023	28/09/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	28/09/2023	28/09/2023
					CT2407	Active implantable	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	28/09/2023	28/09/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	28/09/2023	28/09/2023
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	28/09/2023	28/09/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	28/09/2023	28/09/2023
					CT2727	Active implant devices/ancillaries	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	28/09/2023	28/09/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	28/09/2023	28/09/2023
					CT2750	Implantable cardiac electrical stimulators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	28/09/2023	28/09/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	28/09/2023	28/09/2023
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	28/09/2023	28/09/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	28/09/2023	28/09/2023
					CT2985	Active	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	28/09/2023	28/09/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	28/09/2023	28/09/2023
					CT3001	Implantable metal	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	28/09/2023	28/09/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	28/09/2023	28/09/2023
					CT3002	Implantable synthetic polymer	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	28/09/2023	28/09/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	28/09/2023	28/09/2023
					CT3006	Implantable inorganic material	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	28/09/2023	28/09/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	28/09/2023	28/09/2023
					CT314	Electrically powered	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	28/09/2023	28/09/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	28/09/2023	28/09/2023
					CT321	Long-term surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	28/09/2023	28/09/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	28/09/2023	28/09/2023
					CT449	Generators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	28/09/2023	28/09/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	28/09/2023	28/09/2023
					CT981	Single use	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	28/09/2023	28/09/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	28/09/2023	28/09/2023
					CT983	Surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	28/09/2023	28/09/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	28/09/2023	28/09/2023
14163	2023	All other reportable incidents	35852	Single-chamber implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	05/02/2024	05/02/2024
									A0908	Incorrect, Inadequate or Imprecise Result or Readings	05/02/2024	05/02/2024
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	05/02/2024	05/02/2024
									A0908	Incorrect, Inadequate or Imprecise Result or Readings	05/02/2024	05/02/2024
							F1903	Device Explantation	A070504	Premature Discharge of Battery	05/02/2024	05/02/2024
									A0908	Incorrect, Inadequate or Imprecise Result or Readings	05/02/2024	05/02/2024
14164	2023	All other reportable incidents	37265	Dual-chamber implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
14165	2023	All other reportable incidents	47270	Cardiac resynchronization therapy implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E0119	Loss of Consciousness	A070909	Oversensing		
									A071201	Failure to Convert Rhythm		
									A0713	Defibrillation/Stimulation Problem		
									A072202	Low Impedance		
									A090801	Signal Artifact/Noise		
									A0909	Unexpected Therapeutic Results		
							E011903	Syncope/Fainting	A070909	Oversensing		
									A071201	Failure to Convert Rhythm		
									A0713	Defibrillation/Stimulation Problem		

									A072202	Low Impedance		
									A090801	Signal Artifact/Noise		
									A0909	Unexpected Therapeutic Results		
							E2104	Electric Shock	A070909	Oversensing		
									A071201	Failure to Convert Rhythm		
									A0713	Defibrillation/Stimulation Problem		
									A072202	Low Impedance		
									A090801	Signal Artifact/Noise		
									A0909	Unexpected Therapeutic Results		
							F05	Delay to Treatment/ Therapy	A070909	Oversensing		
									A071201	Failure to Convert Rhythm		
									A0713	Defibrillation/Stimulation Problem		
									A072202	Low Impedance		
									A090801	Signal Artifact/Noise		
									A0909	Unexpected Therapeutic Results		
							F08	Hospitalization or Prolonged Hospitalization	A070909	Oversensing		
									A071201	Failure to Convert Rhythm		
									A0713	Defibrillation/Stimulation Problem		
									A072202	Low Impedance		
									A090801	Signal Artifact/Noise		
									A0909	Unexpected Therapeutic Results		
							F1905	Device Revision or Replacement	A070909	Oversensing		
									A071201	Failure to Convert Rhythm		
									A0713	Defibrillation/Stimulation Problem		
									A072202	Low Impedance		
									A090801	Signal Artifact/Noise		
									A0909	Unexpected Therapeutic Results		
14166	2023		17882	Physiologic-monitoring defibrillation system	CT1684	Cardiac pacemakers and associated devices			A071301	Failure to Deliver Shock/Stimulation		
14167	2023	All other reportable incidents	47265	Dual-chamber implantable pacemaker, rate-responsive	CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A0908	Incorrect, Inadequate or Imprecise Result or Readings	26/09/2023	26/09/2023
							F08	Hospitalization or Prolonged Hospitalization	A0908	Incorrect, Inadequate or Imprecise Result or Readings	26/09/2023	26/09/2023
							F1905	Device Revision or Replacement	A0908	Incorrect, Inadequate or Imprecise Result or Readings	26/09/2023	26/09/2023
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A0908	Incorrect, Inadequate or Imprecise Result or Readings	26/09/2023	26/09/2023
							F08	Hospitalization or Prolonged Hospitalization	A0908	Incorrect, Inadequate or Imprecise Result or Readings	26/09/2023	26/09/2023
							F1905	Device Revision or Replacement	A0908	Incorrect, Inadequate or Imprecise Result or Readings	26/09/2023	26/09/2023
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A0908	Incorrect, Inadequate or Imprecise Result or Readings	26/09/2023	26/09/2023
							F08	Hospitalization or Prolonged Hospitalization	A0908	Incorrect, Inadequate or Imprecise Result or Readings	26/09/2023	26/09/2023
							F1905	Device Revision or Replacement	A0908	Incorrect, Inadequate or Imprecise Result or Readings	26/09/2023	26/09/2023
14168	2023	All other reportable incidents	36241	Pacemaker/defibrillator lead	CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A072201	High Impedance		
							F26	No Health Consequences or Impact	A072201	High Impedance		
14169	2023	All other reportable incidents	17882	Physiologic-monitoring defibrillation system	CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070803	Failure to Power Up		
							F27	Problem Identified During Non-Clinical Procedure	A070803	Failure to Power Up		
14170	2023		47265	Dual-chamber implantable pacemaker, rate-responsive	CT141	Electrodes/Leads and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	09/06/2023	09/06/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	09/06/2023	09/06/2023
					CT145	Cardiovascular devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	09/06/2023	09/06/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	09/06/2023	09/06/2023
					CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	09/06/2023	09/06/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	09/06/2023	09/06/2023
					CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	09/06/2023	09/06/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	09/06/2023	09/06/2023
					CT233	Surgical	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	09/06/2023	09/06/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	09/06/2023	09/06/2023
					CT2406	Implantable	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	09/06/2023	09/06/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	09/06/2023	09/06/2023
					CT2407	Active implantable	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	09/06/2023	09/06/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	09/06/2023	09/06/2023
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	09/06/2023	09/06/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	09/06/2023	09/06/2023
					CT2727	Active implant devices/ancillaries	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	09/06/2023	09/06/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	09/06/2023	09/06/2023
					CT2750	Implantable cardiac electrical stimulators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	09/06/2023	09/06/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	09/06/2023	09/06/2023
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	09/06/2023	09/06/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	09/06/2023	09/06/2023
					CT2985	Active	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	09/06/2023	09/06/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	09/06/2023	09/06/2023
					CT3001	Implantable metal	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	09/06/2023	09/06/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	09/06/2023	09/06/2023
					CT3002	Implantable synthetic polymer	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	09/06/2023	09/06/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	09/06/2023	09/06/2023
					CT3006	Implantable inorganic material	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	09/06/2023	09/06/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	09/06/2023	09/06/2023
					CT314	Electrically powered	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	09/06/2023	09/06/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	09/06/2023	09/06/2023
					CT321	Long-term surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	09/06/2023	09/06/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	09/06/2023	09/06/2023
					CT449	Generators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	09/06/2023	09/06/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	09/06/2023	09/06/2023
					CT981	Single use	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	09/06/2023	09/06/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	09/06/2023	09/06/2023
					CT983	Surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	09/06/2023	09/06/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	09/06/2023	09/06/2023

14171	2023		47265	Dual-chamber implantable pacemaker, rate-responsive	CT141	Electrodes/Leads and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	19/09/2023	19/09/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	19/09/2023	19/09/2023
					CT145	Cardiovascular devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	19/09/2023	19/09/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	19/09/2023	19/09/2023
					CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	19/09/2023	19/09/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	19/09/2023	19/09/2023
					CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	19/09/2023	19/09/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	19/09/2023	19/09/2023
					CT233	Surgical	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	19/09/2023	19/09/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	19/09/2023	19/09/2023
					CT2406	Implantable	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	19/09/2023	19/09/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	19/09/2023	19/09/2023
					CT2407	Active implantable	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	19/09/2023	19/09/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	19/09/2023	19/09/2023
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	19/09/2023	19/09/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	19/09/2023	19/09/2023
					CT2727	Active implant devices/ancillaries	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	19/09/2023	19/09/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	19/09/2023	19/09/2023
					CT2750	Implantable cardiac electrical stimulators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	19/09/2023	19/09/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	19/09/2023	19/09/2023
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	19/09/2023	19/09/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	19/09/2023	19/09/2023
					CT2985	Active	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	19/09/2023	19/09/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	19/09/2023	19/09/2023
					CT3001	Implantable metal	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	19/09/2023	19/09/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	19/09/2023	19/09/2023
					CT3002	Implantable synthetic polymer	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	19/09/2023	19/09/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	19/09/2023	19/09/2023
					CT3006	Implantable inorganic material	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	19/09/2023	19/09/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	19/09/2023	19/09/2023
					CT314	Electrically powered	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	19/09/2023	19/09/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	19/09/2023	19/09/2023
					CT321	Long-term surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	19/09/2023	19/09/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	19/09/2023	19/09/2023
					CT449	Generators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	19/09/2023	19/09/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	19/09/2023	19/09/2023
					CT981	Single use	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	19/09/2023	19/09/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	19/09/2023	19/09/2023
					CT983	Surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	19/09/2023	19/09/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	19/09/2023	19/09/2023
14172	2023	All other reportable incidents	17882	Physiologic-monitoring defibrillation system	CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A090406	Therapeutic or Diagnostic Output Failure		
							F27	Problem Identified During Non-Clinical Procedure	A090406	Therapeutic or Diagnostic Output Failure		
14173	2023		37265	Dual-chamber implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A160601	Premature Elective Replacement Indicator	22/09/2023	
							F26	No Health Consequences or Impact	A160601	Premature Elective Replacement Indicator	22/09/2023	
14174	2023	All other reportable incidents	36241	Pacemaker/defibrillator lead	CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070101	Failure to Capture		
									A072201	High Impedance		
							F26	No Health Consequences or Impact	A070101	Failure to Capture		
									A072201	High Impedance		
14175	2023	All other reportable incidents	47267	Single-chamber implantable pacemaker, rate-responsive	CT141	Electrodes/Leads and associated devices	E2103	Device Overstimulation of Tissue	A0908	Incorrect, Inadequate or Imprecise Result or Readings	28/09/2023	28/09/2023
							F08	Hospitalization or Prolonged Hospitalization	A0908	Incorrect, Inadequate or Imprecise Result or Readings	28/09/2023	28/09/2023
							F1905	Device Revision or Replacement	A0908	Incorrect, Inadequate or Imprecise Result or Readings	28/09/2023	28/09/2023
					CT145	Cardiovascular devices	E2103	Device Overstimulation of Tissue	A0908	Incorrect, Inadequate or Imprecise Result or Readings	28/09/2023	28/09/2023
							F08	Hospitalization or Prolonged Hospitalization	A0908	Incorrect, Inadequate or Imprecise Result or Readings	28/09/2023	28/09/2023
							F1905	Device Revision or Replacement	A0908	Incorrect, Inadequate or Imprecise Result or Readings	28/09/2023	28/09/2023
					CT1684	Cardiac pacemakers and associated devices	E2103	Device Overstimulation of Tissue	A0908	Incorrect, Inadequate or Imprecise Result or Readings	28/09/2023	28/09/2023
							F08	Hospitalization or Prolonged Hospitalization	A0908	Incorrect, Inadequate or Imprecise Result or Readings	28/09/2023	28/09/2023
							F1905	Device Revision or Replacement	A0908	Incorrect, Inadequate or Imprecise Result or Readings	28/09/2023	28/09/2023
					CT1685	Cardiac pacemakers	E2103	Device Overstimulation of Tissue	A0908	Incorrect, Inadequate or Imprecise Result or Readings	28/09/2023	28/09/2023
							F08	Hospitalization or Prolonged Hospitalization	A0908	Incorrect, Inadequate or Imprecise Result or Readings	28/09/2023	28/09/2023
							F1905	Device Revision or Replacement	A0908	Incorrect, Inadequate or Imprecise Result or Readings	28/09/2023	28/09/2023
					CT233	Surgical	E2103	Device Overstimulation of Tissue	A0908	Incorrect, Inadequate or Imprecise Result or Readings	28/09/2023	28/09/2023
							F08	Hospitalization or Prolonged Hospitalization	A0908	Incorrect, Inadequate or Imprecise Result or Readings	28/09/2023	28/09/2023
							F1905	Device Revision or Replacement	A0908	Incorrect, Inadequate or Imprecise Result or Readings	28/09/2023	28/09/2023
					CT2406	Implantable	E2103	Device Overstimulation of Tissue	A0908	Incorrect, Inadequate or Imprecise Result or Readings	28/09/2023	28/09/2023
							F08	Hospitalization or Prolonged Hospitalization	A0908	Incorrect, Inadequate or Imprecise Result or Readings	28/09/2023	28/09/2023
							F1905	Device Revision or Replacement	A0908	Incorrect, Inadequate or Imprecise Result or Readings	28/09/2023	28/09/2023
					CT2407	Active implantable	E2103	Device Overstimulation of Tissue	A0908	Incorrect, Inadequate or Imprecise Result or Readings	28/09/2023	28/09/2023
							F08	Hospitalization or Prolonged Hospitalization	A0908	Incorrect, Inadequate or Imprecise Result or Readings	28/09/2023	28/09/2023
							F1905	Device Revision or Replacement	A0908	Incorrect, Inadequate or Imprecise Result or Readings	28/09/2023	28/09/2023
					CT257	Implantable cardiac pacemakers	E2103	Device Overstimulation of Tissue	A0908	Incorrect, Inadequate or Imprecise Result or Readings	28/09/2023	28/09/2023

							F08	Hospitalization or Prolonged Hospitalization	A0908	Incorrect, Inadequate or Imprecise Result or Readings	28/09/2023	28/09/2023
							F1905	Device Revision or Replacement	A0908	Incorrect, Inadequate or Imprecise Result or Readings	28/09/2023	28/09/2023
					CT2727	Active implant devices/ancillaries	E2103	Device Overstimulation of Tissue	A0908	Incorrect, Inadequate or Imprecise Result or Readings	28/09/2023	28/09/2023
							F08	Hospitalization or Prolonged Hospitalization	A0908	Incorrect, Inadequate or Imprecise Result or Readings	28/09/2023	28/09/2023
							F1905	Device Revision or Replacement	A0908	Incorrect, Inadequate or Imprecise Result or Readings	28/09/2023	28/09/2023
					CT2750	Implantable cardiac electrical stimulators and associated devices	E2103	Device Overstimulation of Tissue	A0908	Incorrect, Inadequate or Imprecise Result or Readings	28/09/2023	28/09/2023
							F08	Hospitalization or Prolonged Hospitalization	A0908	Incorrect, Inadequate or Imprecise Result or Readings	28/09/2023	28/09/2023
							F1905	Device Revision or Replacement	A0908	Incorrect, Inadequate or Imprecise Result or Readings	28/09/2023	28/09/2023
					CT2866	Implantable cardiac electrical stimulators	E2103	Device Overstimulation of Tissue	A0908	Incorrect, Inadequate or Imprecise Result or Readings	28/09/2023	28/09/2023
							F08	Hospitalization or Prolonged Hospitalization	A0908	Incorrect, Inadequate or Imprecise Result or Readings	28/09/2023	28/09/2023
							F1905	Device Revision or Replacement	A0908	Incorrect, Inadequate or Imprecise Result or Readings	28/09/2023	28/09/2023
					CT2985	Active	E2103	Device Overstimulation of Tissue	A0908	Incorrect, Inadequate or Imprecise Result or Readings	28/09/2023	28/09/2023
							F08	Hospitalization or Prolonged Hospitalization	A0908	Incorrect, Inadequate or Imprecise Result or Readings	28/09/2023	28/09/2023
							F1905	Device Revision or Replacement	A0908	Incorrect, Inadequate or Imprecise Result or Readings	28/09/2023	28/09/2023
					CT3001	Implantable metal	E2103	Device Overstimulation of Tissue	A0908	Incorrect, Inadequate or Imprecise Result or Readings	28/09/2023	28/09/2023
							F08	Hospitalization or Prolonged Hospitalization	A0908	Incorrect, Inadequate or Imprecise Result or Readings	28/09/2023	28/09/2023
							F1905	Device Revision or Replacement	A0908	Incorrect, Inadequate or Imprecise Result or Readings	28/09/2023	28/09/2023
					CT3002	Implantable synthetic polymer	E2103	Device Overstimulation of Tissue	A0908	Incorrect, Inadequate or Imprecise Result or Readings	28/09/2023	28/09/2023
							F08	Hospitalization or Prolonged Hospitalization	A0908	Incorrect, Inadequate or Imprecise Result or Readings	28/09/2023	28/09/2023
							F1905	Device Revision or Replacement	A0908	Incorrect, Inadequate or Imprecise Result or Readings	28/09/2023	28/09/2023
					CT3006	Implantable inorganic material	E2103	Device Overstimulation of Tissue	A0908	Incorrect, Inadequate or Imprecise Result or Readings	28/09/2023	28/09/2023
							F08	Hospitalization or Prolonged Hospitalization	A0908	Incorrect, Inadequate or Imprecise Result or Readings	28/09/2023	28/09/2023
							F1905	Device Revision or Replacement	A0908	Incorrect, Inadequate or Imprecise Result or Readings	28/09/2023	28/09/2023
					CT314	Electrically powered	E2103	Device Overstimulation of Tissue	A0908	Incorrect, Inadequate or Imprecise Result or Readings	28/09/2023	28/09/2023
							F08	Hospitalization or Prolonged Hospitalization	A0908	Incorrect, Inadequate or Imprecise Result or Readings	28/09/2023	28/09/2023
							F1905	Device Revision or Replacement	A0908	Incorrect, Inadequate or Imprecise Result or Readings	28/09/2023	28/09/2023
					CT321	Long-term surgical invasive	E2103	Device Overstimulation of Tissue	A0908	Incorrect, Inadequate or Imprecise Result or Readings	28/09/2023	28/09/2023
							F08	Hospitalization or Prolonged Hospitalization	A0908	Incorrect, Inadequate or Imprecise Result or Readings	28/09/2023	28/09/2023
							F1905	Device Revision or Replacement	A0908	Incorrect, Inadequate or Imprecise Result or Readings	28/09/2023	28/09/2023
					CT449	Generators and associated devices	E2103	Device Overstimulation of Tissue	A0908	Incorrect, Inadequate or Imprecise Result or Readings	28/09/2023	28/09/2023
							F08	Hospitalization or Prolonged Hospitalization	A0908	Incorrect, Inadequate or Imprecise Result or Readings	28/09/2023	28/09/2023
							F1905	Device Revision or Replacement	A0908	Incorrect, Inadequate or Imprecise Result or Readings	28/09/2023	28/09/2023
					CT981	Single use	E2103	Device Overstimulation of Tissue	A0908	Incorrect, Inadequate or Imprecise Result or Readings	28/09/2023	28/09/2023
							F08	Hospitalization or Prolonged Hospitalization	A0908	Incorrect, Inadequate or Imprecise Result or Readings	28/09/2023	28/09/2023
							F1905	Device Revision or Replacement	A0908	Incorrect, Inadequate or Imprecise Result or Readings	28/09/2023	28/09/2023
					CT983	Surgical invasive	E2103	Device Overstimulation of Tissue	A0908	Incorrect, Inadequate or Imprecise Result or Readings	28/09/2023	28/09/2023
							F08	Hospitalization or Prolonged Hospitalization	A0908	Incorrect, Inadequate or Imprecise Result or Readings	28/09/2023	28/09/2023
							F1905	Device Revision or Replacement	A0908	Incorrect, Inadequate or Imprecise Result or Readings	28/09/2023	28/09/2023
14176	2023	All other reportable incidents	36241	Pacemaker/defibrillator lead	CT1684	Cardiac pacemakers and associated devices	E0604	Cardiac Perforation	A070102	High Capture Threshold		
									A070909	Oversensing		
									A090801	Signal Artifact/Noise		
									A1301	Failure to Read Input Signal		
							F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A070102	High Capture Threshold		
									A070909	Oversensing		
									A090801	Signal Artifact/Noise		
									A1301	Failure to Read Input Signal		
							F1203	Life Threatening Illness or Injury	A070102	High Capture Threshold		
									A070909	Oversensing		
									A090801	Signal Artifact/Noise		
									A1301	Failure to Read Input Signal		
14177	2023	All other reportable incidents	47263	Cardiac resynchronization therapy implantable pacemaker	CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070909	Oversensing	05/10/2023	05/10/2023
									A0712	Pacing Problem	05/10/2023	05/10/2023
									A0908	Incorrect, Inadequate or Imprecise Result or Readings	05/10/2023	05/10/2023
									A090801	Signal Artifact/Noise	05/10/2023	05/10/2023
							F08	Hospitalization or Prolonged Hospitalization	A070909	Oversensing	05/10/2023	05/10/2023
									A0712	Pacing Problem	05/10/2023	05/10/2023
									A0908	Incorrect, Inadequate or Imprecise Result or Readings	05/10/2023	05/10/2023

									A090801	Signal Artifact/Noise	05/10/2023	05/10/2023
							F1001	Absence of Treatment	A070909	Oversensing	05/10/2023	05/10/2023
									A0712	Pacing Problem	05/10/2023	05/10/2023
									A0908	Incorrect, Inadequate or Imprecise Result or Readings	05/10/2023	05/10/2023
									A090801	Signal Artifact/Noise	05/10/2023	05/10/2023
							F1905	Device Revision or Replacement	A070909	Oversensing	05/10/2023	05/10/2023
									A0712	Pacing Problem	05/10/2023	05/10/2023
									A0908	Incorrect, Inadequate or Imprecise Result or Readings	05/10/2023	05/10/2023
									A090801	Signal Artifact/Noise	05/10/2023	05/10/2023
					CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A070909	Oversensing	05/10/2023	05/10/2023
									A0712	Pacing Problem	05/10/2023	05/10/2023
									A0908	Incorrect, Inadequate or Imprecise Result or Readings	05/10/2023	05/10/2023
									A090801	Signal Artifact/Noise	05/10/2023	05/10/2023
							F08	Hospitalization or Prolonged Hospitalization	A070909	Oversensing	05/10/2023	05/10/2023
									A0712	Pacing Problem	05/10/2023	05/10/2023
									A0908	Incorrect, Inadequate or Imprecise Result or Readings	05/10/2023	05/10/2023
									A090801	Signal Artifact/Noise	05/10/2023	05/10/2023
							F1001	Absence of Treatment	A070909	Oversensing	05/10/2023	05/10/2023
									A0712	Pacing Problem	05/10/2023	05/10/2023
									A0908	Incorrect, Inadequate or Imprecise Result or Readings	05/10/2023	05/10/2023
									A090801	Signal Artifact/Noise	05/10/2023	05/10/2023
							F1905	Device Revision or Replacement	A070909	Oversensing	05/10/2023	05/10/2023
									A0712	Pacing Problem	05/10/2023	05/10/2023
									A0908	Incorrect, Inadequate or Imprecise Result or Readings	05/10/2023	05/10/2023
									A090801	Signal Artifact/Noise	05/10/2023	05/10/2023
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A070909	Oversensing	05/10/2023	05/10/2023
									A0712	Pacing Problem	05/10/2023	05/10/2023
									A0908	Incorrect, Inadequate or Imprecise Result or Readings	05/10/2023	05/10/2023
									A090801	Signal Artifact/Noise	05/10/2023	05/10/2023
							F08	Hospitalization or Prolonged Hospitalization	A070909	Oversensing	05/10/2023	05/10/2023
									A0712	Pacing Problem	05/10/2023	05/10/2023
									A0908	Incorrect, Inadequate or Imprecise Result or Readings	05/10/2023	05/10/2023
									A090801	Signal Artifact/Noise	05/10/2023	05/10/2023
							F1001	Absence of Treatment	A070909	Oversensing	05/10/2023	05/10/2023
									A0712	Pacing Problem	05/10/2023	05/10/2023
									A0908	Incorrect, Inadequate or Imprecise Result or Readings	05/10/2023	05/10/2023
									A090801	Signal Artifact/Noise	05/10/2023	05/10/2023
							F1905	Device Revision or Replacement	A070909	Oversensing	05/10/2023	05/10/2023
									A0712	Pacing Problem	05/10/2023	05/10/2023
									A0908	Incorrect, Inadequate or Imprecise Result or Readings	05/10/2023	05/10/2023
									A090801	Signal Artifact/Noise	05/10/2023	05/10/2023
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A070909	Oversensing	05/10/2023	05/10/2023
									A0712	Pacing Problem	05/10/2023	05/10/2023
									A0908	Incorrect, Inadequate or Imprecise Result or Readings	05/10/2023	05/10/2023
									A090801	Signal Artifact/Noise	05/10/2023	05/10/2023
							F08	Hospitalization or Prolonged Hospitalization	A070909	Oversensing	05/10/2023	05/10/2023
									A0712	Pacing Problem	05/10/2023	05/10/2023
									A0908	Incorrect, Inadequate or Imprecise Result or Readings	05/10/2023	05/10/2023
									A090801	Signal Artifact/Noise	05/10/2023	05/10/2023
							F1001	Absence of Treatment	A070909	Oversensing	05/10/2023	05/10/2023
									A0712	Pacing Problem	05/10/2023	05/10/2023
									A0908	Incorrect, Inadequate or Imprecise Result or Readings	05/10/2023	05/10/2023
									A090801	Signal Artifact/Noise	05/10/2023	05/10/2023
							F1905	Device Revision or Replacement	A070909	Oversensing	05/10/2023	05/10/2023
									A0712	Pacing Problem	05/10/2023	05/10/2023
									A0908	Incorrect, Inadequate or Imprecise Result or Readings	05/10/2023	05/10/2023
									A090801	Signal Artifact/Noise	05/10/2023	05/10/2023
14178	2023	All other reportable incidents	47265	Dual-chamber implantable pacemaker, rate-responsive	CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
14179	2023	All other reportable incidents	47265	Dual-chamber implantable pacemaker, rate-responsive	CT141	Electrodes/Leads and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070102	High Capture Threshold		
									A072202	Low Impedance		
									A090801	Signal Artifact/Noise		
							F26	No Health Consequences or Impact	A070102	High Capture Threshold		
									A072202	Low Impedance		
									A090801	Signal Artifact/Noise		
					CT145	Cardiovascular devices	E2403	No Clinical Signs, Symptoms or Conditions	A070102	High Capture Threshold		
									A072202	Low Impedance		

								A090801	Signal Artifact/Noise		
						F26	No Health Consequences or Impact	A070102	High Capture Threshold		
								A072202	Low Impedance		
								A090801	Signal Artifact/Noise		
					CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070102	High Capture Threshold		
									A072202	Low Impedance		
									A090801	Signal Artifact/Noise		
							F26	No Health Consequences or Impact	A070102	High Capture Threshold		
									A072202	Low Impedance		
									A090801	Signal Artifact/Noise		
					CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A070102	High Capture Threshold		
									A072202	Low Impedance		
									A090801	Signal Artifact/Noise		
							F26	No Health Consequences or Impact	A070102	High Capture Threshold		
									A072202	Low Impedance		
									A090801	Signal Artifact/Noise		
					CT233	Surgical	E2403	No Clinical Signs, Symptoms or Conditions	A070102	High Capture Threshold		
									A072202	Low Impedance		
									A090801	Signal Artifact/Noise		
							F26	No Health Consequences or Impact	A070102	High Capture Threshold		
									A072202	Low Impedance		
									A090801	Signal Artifact/Noise		
					CT2406	Implantable	E2403	No Clinical Signs, Symptoms or Conditions	A070102	High Capture Threshold		
									A072202	Low Impedance		
									A090801	Signal Artifact/Noise		
							F26	No Health Consequences or Impact	A070102	High Capture Threshold		
									A072202	Low Impedance		
									A090801	Signal Artifact/Noise		
					CT2407	Active implantable	E2403	No Clinical Signs, Symptoms or Conditions	A070102	High Capture Threshold		
									A072202	Low Impedance		
									A090801	Signal Artifact/Noise		
							F26	No Health Consequences or Impact	A070102	High Capture Threshold		
									A072202	Low Impedance		
									A090801	Signal Artifact/Noise		
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A070102	High Capture Threshold		
									A072202	Low Impedance		
									A090801	Signal Artifact/Noise		
							F26	No Health Consequences or Impact	A070102	High Capture Threshold		
									A072202	Low Impedance		
									A090801	Signal Artifact/Noise		
					CT2727	Active implant devices/ancillaries	E2403	No Clinical Signs, Symptoms or Conditions	A070102	High Capture Threshold		
									A072202	Low Impedance		
									A090801	Signal Artifact/Noise		
							F26	No Health Consequences or Impact	A070102	High Capture Threshold		
									A072202	Low Impedance		
									A090801	Signal Artifact/Noise		
					CT2750	Implantable cardiac electrical stimulators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070102	High Capture Threshold		
									A072202	Low Impedance		
									A090801	Signal Artifact/Noise		
							F26	No Health Consequences or Impact	A070102	High Capture Threshold		
									A072202	Low Impedance		
									A090801	Signal Artifact/Noise		
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A070102	High Capture Threshold		
									A072202	Low Impedance		
									A090801	Signal Artifact/Noise		
							F26	No Health Consequences or Impact	A070102	High Capture Threshold		
									A072202	Low Impedance		
									A090801	Signal Artifact/Noise		
					CT2985	Active	E2403	No Clinical Signs, Symptoms or Conditions	A070102	High Capture Threshold		
									A072202	Low Impedance		
									A090801	Signal Artifact/Noise		
							F26	No Health Consequences or Impact	A070102	High Capture Threshold		
									A072202	Low Impedance		
									A090801	Signal Artifact/Noise		
					CT3001	Implantable metal	E2403	No Clinical Signs, Symptoms or Conditions	A070102	High Capture Threshold		
									A072202	Low Impedance		
									A090801	Signal Artifact/Noise		
							F26	No Health Consequences or Impact	A070102	High Capture Threshold		
									A072202	Low Impedance		
									A090801	Signal Artifact/Noise		
					CT3002	Implantable synthetic polymer	E2403	No Clinical Signs, Symptoms or Conditions	A070102	High Capture Threshold		
									A072202	Low Impedance		
									A090801	Signal Artifact/Noise		
							F26	No Health Consequences or Impact	A070102	High Capture Threshold		
									A072202	Low Impedance		
									A090801	Signal Artifact/Noise		
					CT3006	Implantable inorganic material	E2403	No Clinical Signs, Symptoms or Conditions	A070102	High Capture Threshold		
									A072202	Low Impedance		
									A090801	Signal Artifact/Noise		
							F26	No Health Consequences or Impact	A070102	High Capture Threshold		
									A072202	Low Impedance		
									A090801	Signal Artifact/Noise		
					CT314	Electrically powered	E2403	No Clinical Signs, Symptoms or Conditions	A070102	High Capture Threshold		
									A072202	Low Impedance		
									A090801	Signal Artifact/Noise		
							F26	No Health Consequences or Impact	A070102	High Capture Threshold		
									A072202	Low Impedance		
									A090801	Signal Artifact/Noise		

					CT321	Long-term surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A070102	High Capture Threshold		
									A072202	Low Impedance		
									A090801	Signal Artifact/Noise		
							F26	No Health Consequences or Impact	A070102	High Capture Threshold		
									A072202	Low Impedance		
									A090801	Signal Artifact/Noise		
					CT449	Generators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070102	High Capture Threshold		
									A072202	Low Impedance		
									A090801	Signal Artifact/Noise		
							F26	No Health Consequences or Impact	A070102	High Capture Threshold		
									A072202	Low Impedance		
									A090801	Signal Artifact/Noise		
					CT981	Single use	E2403	No Clinical Signs, Symptoms or Conditions	A070102	High Capture Threshold		
									A072202	Low Impedance		
									A090801	Signal Artifact/Noise		
							F26	No Health Consequences or Impact	A070102	High Capture Threshold		
									A072202	Low Impedance		
									A090801	Signal Artifact/Noise		
					CT983	Surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A070102	High Capture Threshold		
									A072202	Low Impedance		
									A090801	Signal Artifact/Noise		
							F26	No Health Consequences or Impact	A070102	High Capture Threshold		
									A072202	Low Impedance		
									A090801	Signal Artifact/Noise		
14180	2023	All other reportable incidents	47265	Dual-chamber implantable pacemaker, rate-responsive	CT141	Electrodes/Leads and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070909	Oversensing		
									A071204	No Pacing		
							F1901	Additional Surgery	A070909	Oversensing		
									A071204	No Pacing		
					CT145	Cardiovascular devices	E2403	No Clinical Signs, Symptoms or Conditions	A070909	Oversensing		
									A071204	No Pacing		
							F1901	Additional Surgery	A070909	Oversensing		
									A071204	No Pacing		
					CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070909	Oversensing		
									A071204	No Pacing		
							F1901	Additional Surgery	A070909	Oversensing		
									A071204	No Pacing		
					CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A070909	Oversensing		
									A071204	No Pacing		
							F1901	Additional Surgery	A070909	Oversensing		
									A071204	No Pacing		
					CT233	Surgical	E2403	No Clinical Signs, Symptoms or Conditions	A070909	Oversensing		
									A071204	No Pacing		
							F1901	Additional Surgery	A070909	Oversensing		
									A071204	No Pacing		
					CT2406	Implantable	E2403	No Clinical Signs, Symptoms or Conditions	A070909	Oversensing		
									A071204	No Pacing		
							F1901	Additional Surgery	A070909	Oversensing		
									A071204	No Pacing		
					CT2407	Active implantable	E2403	No Clinical Signs, Symptoms or Conditions	A070909	Oversensing		
									A071204	No Pacing		
							F1901	Additional Surgery	A070909	Oversensing		
									A071204	No Pacing		
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A070909	Oversensing		
									A071204	No Pacing		
							F1901	Additional Surgery	A070909	Oversensing		
									A071204	No Pacing		
					CT2727	Active implant devices/ancillaries	E2403	No Clinical Signs, Symptoms or Conditions	A070909	Oversensing		
									A071204	No Pacing		
							F1901	Additional Surgery	A070909	Oversensing		
									A071204	No Pacing		
					CT2750	Implantable cardiac electrical stimulators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070909	Oversensing		
									A071204	No Pacing		
							F1901	Additional Surgery	A070909	Oversensing		
									A071204	No Pacing		
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A070909	Oversensing		
									A071204	No Pacing		
							F1901	Additional Surgery	A070909	Oversensing		
									A071204	No Pacing		
					CT2985	Active	E2403	No Clinical Signs, Symptoms or Conditions	A070909	Oversensing		
									A071204	No Pacing		
							F1901	Additional Surgery	A070909	Oversensing		
									A071204	No Pacing		
					CT3001	Implantable metal	E2403	No Clinical Signs, Symptoms or Conditions	A070909	Oversensing		
									A071204	No Pacing		
							F1901	Additional Surgery	A070909	Oversensing		
									A071204	No Pacing		
					CT3002	Implantable synthetic polymer	E2403	No Clinical Signs, Symptoms or Conditions	A070909	Oversensing		
									A071204	No Pacing		
							F1901	Additional Surgery	A070909	Oversensing		
									A071204	No Pacing		
					CT3006	Implantable inorganic material	E2403	No Clinical Signs, Symptoms or Conditions	A070909	Oversensing		
									A071204	No Pacing		
							F1901	Additional Surgery	A070909	Oversensing		
									A071204	No Pacing		
					CT314	Electrically powered	E2403	No Clinical Signs, Symptoms or Conditions	A070909	Oversensing		
									A071204	No Pacing		
							F1901	Additional Surgery	A070909	Oversensing		
									A071204	No Pacing		

					CT321	Long-term surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A070909	Oversensing		
									A071204	No Pacing		
							F1901	Additional Surgery	A070909	Oversensing		
									A071204	No Pacing		
					CT449	Generators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070909	Oversensing		
									A071204	No Pacing		
							F1901	Additional Surgery	A070909	Oversensing		
									A071204	No Pacing		
					CT981	Single use	E2403	No Clinical Signs, Symptoms or Conditions	A070909	Oversensing		
									A071204	No Pacing		
							F1901	Additional Surgery	A070909	Oversensing		
									A071204	No Pacing		
					CT983	Surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A070909	Oversensing		
									A071204	No Pacing		
							F1901	Additional Surgery	A070909	Oversensing		
									A071204	No Pacing		
14181	2023	All other reportable incidents	17882	Physiologic-monitoring defibrillation system	CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A0713	Defibrillation/Stimulation Problem		
							F27	Problem Identified During Non-Clinical Procedure	A0713	Defibrillation/Stimulation Problem		
14182	2023	All other reportable incidents	47270	Cardiac resynchronization therapy implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A090811	Unable to Obtain Readings		
							F26	No Health Consequences or Impact	A090811	Unable to Obtain Readings		
14183	2023	All other reportable incidents	17882	Physiologic-monitoring defibrillation system	CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A1603	Failure of Device to Self-Test		
							F27	Problem Identified During Non-Clinical Procedure	A1603	Failure of Device to Self-Test		
14184	2023	All other reportable incidents	17882	Physiologic-monitoring defibrillation system	CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A1603	Failure of Device to Self-Test		
							F27	Problem Identified During Non-Clinical Procedure	A1603	Failure of Device to Self-Test		
14185	2023	All other reportable incidents	17882	Physiologic-monitoring defibrillation system	CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A090811	Unable to Obtain Readings		
							F27	Problem Identified During Non-Clinical Procedure	A090811	Unable to Obtain Readings		
14186	2023	All other reportable incidents	35852	Single-chamber implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	13/10/2023	13/10/2023
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	13/10/2023	13/10/2023
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	13/10/2023	13/10/2023
14187	2023	All other reportable incidents	47265	Dual-chamber implantable pacemaker, rate-responsive	CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
14188	2023	All other reportable incidents	47263	Cardiac resynchronization therapy implantable pacemaker	CT1684	Cardiac pacemakers and associated devices	E011903	Syncope/Fainting	A070101	Failure to Capture	04/10/2023	04/10/2023
									A0908	Incorrect, Inadequate or Imprecise Result or Readings	04/10/2023	04/10/2023
							E060104	Bradycardia	A070101	Failure to Capture	04/10/2023	04/10/2023
									A0908	Incorrect, Inadequate or Imprecise Result or Readings	04/10/2023	04/10/2023
							F08	Hospitalization or Prolonged Hospitalization	A070101	Failure to Capture	04/10/2023	04/10/2023
									A0908	Incorrect, Inadequate or Imprecise Result or Readings	04/10/2023	04/10/2023
							F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A070101	Failure to Capture	04/10/2023	04/10/2023
									A0908	Incorrect, Inadequate or Imprecise Result or Readings	04/10/2023	04/10/2023
							F1905	Device Revision or Replacement	A070101	Failure to Capture	04/10/2023	04/10/2023
									A0908	Incorrect, Inadequate or Imprecise Result or Readings	04/10/2023	04/10/2023
					CT1685	Cardiac pacemakers	E011903	Syncope/Fainting	A070101	Failure to Capture	04/10/2023	04/10/2023
									A0908	Incorrect, Inadequate or Imprecise Result or Readings	04/10/2023	04/10/2023
							E060104	Bradycardia	A070101	Failure to Capture	04/10/2023	04/10/2023
									A0908	Incorrect, Inadequate or Imprecise Result or Readings	04/10/2023	04/10/2023
							F08	Hospitalization or Prolonged Hospitalization	A070101	Failure to Capture	04/10/2023	04/10/2023
									A0908	Incorrect, Inadequate or Imprecise Result or Readings	04/10/2023	04/10/2023
							F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A070101	Failure to Capture	04/10/2023	04/10/2023
									A0908	Incorrect, Inadequate or Imprecise Result or Readings	04/10/2023	04/10/2023
							F1905	Device Revision or Replacement	A070101	Failure to Capture	04/10/2023	04/10/2023
									A0908	Incorrect, Inadequate or Imprecise Result or Readings	04/10/2023	04/10/2023
					CT257	Implantable cardiac pacemakers	E011903	Syncope/Fainting	A070101	Failure to Capture	04/10/2023	04/10/2023
									A0908	Incorrect, Inadequate or Imprecise Result or Readings	04/10/2023	04/10/2023
							E060104	Bradycardia	A070101	Failure to Capture	04/10/2023	04/10/2023
									A0908	Incorrect, Inadequate or Imprecise Result or Readings	04/10/2023	04/10/2023
							F08	Hospitalization or Prolonged Hospitalization	A070101	Failure to Capture	04/10/2023	04/10/2023
									A0908	Incorrect, Inadequate or Imprecise Result or Readings	04/10/2023	04/10/2023
							F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A070101	Failure to Capture	04/10/2023	04/10/2023
									A0908	Incorrect, Inadequate or Imprecise Result or Readings	04/10/2023	04/10/2023
							F1905	Device Revision or Replacement	A070101	Failure to Capture	04/10/2023	04/10/2023
									A0908	Incorrect, Inadequate or Imprecise Result or Readings	04/10/2023	04/10/2023
					CT2866	Implantable cardiac electrical stimulators	E011903	Syncope/Fainting	A070101	Failure to Capture	04/10/2023	04/10/2023
									A0908	Incorrect, Inadequate or Imprecise Result or Readings	04/10/2023	04/10/2023
							E060104	Bradycardia	A070101	Failure to Capture	04/10/2023	04/10/2023
									A0908	Incorrect, Inadequate or Imprecise Result or Readings	04/10/2023	04/10/2023

							F08	Hospitalization or Prolonged Hospitalization	A070101	Failure to Capture	04/10/2023	04/10/2023
									A0908	Incorrect, Inadequate or Imprecise Result or Readings	04/10/2023	04/10/2023
							F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A070101	Failure to Capture	04/10/2023	04/10/2023
									A0908	Incorrect, Inadequate or Imprecise Result or Readings	04/10/2023	04/10/2023
							F1905	Device Revision or Replacement	A070101	Failure to Capture	04/10/2023	04/10/2023
									A0908	Incorrect, Inadequate or Imprecise Result or Readings	04/10/2023	04/10/2023
14189	2023	All other reportable incidents	47265	Dual-chamber implantable pacemaker, rate-responsive	CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
14190	2023	All other reportable incidents	17882	Physiologic-monitoring defibrillation system	CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070908	Failure to Sense		
									A0719	Unexpected Shutdown		
							F26	No Health Consequences or Impact	A070908	Failure to Sense		
									A0719	Unexpected Shutdown		
14191	2023	All other reportable incidents	47265	Dual-chamber implantable pacemaker, rate-responsive	CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem	18/10/2022	18/10/2022
									A070504	Premature Discharge of Battery	18/10/2022	18/10/2022
							F08	Hospitalization or Prolonged Hospitalization	A0705	Battery Problem	18/10/2022	18/10/2022
									A070504	Premature Discharge of Battery	18/10/2022	18/10/2022
							F1905	Device Revision or Replacement	A0705	Battery Problem	18/10/2022	18/10/2022
									A070504	Premature Discharge of Battery	18/10/2022	18/10/2022
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem	18/10/2022	18/10/2022
									A070504	Premature Discharge of Battery	18/10/2022	18/10/2022
							F08	Hospitalization or Prolonged Hospitalization	A0705	Battery Problem	18/10/2022	18/10/2022
									A070504	Premature Discharge of Battery	18/10/2022	18/10/2022
							F1905	Device Revision or Replacement	A0705	Battery Problem	18/10/2022	18/10/2022
									A070504	Premature Discharge of Battery	18/10/2022	18/10/2022
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem	18/10/2022	18/10/2022
									A070504	Premature Discharge of Battery	18/10/2022	18/10/2022
							F08	Hospitalization or Prolonged Hospitalization	A0705	Battery Problem	18/10/2022	18/10/2022
									A070504	Premature Discharge of Battery	18/10/2022	18/10/2022
							F1905	Device Revision or Replacement	A0705	Battery Problem	18/10/2022	18/10/2022
									A070504	Premature Discharge of Battery	18/10/2022	18/10/2022
14192	2023	All other reportable incidents	47265	Dual-chamber implantable pacemaker, rate-responsive	CT141	Electrodes/Leads and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	29/08/2023	29/08/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	29/08/2023	29/08/2023
					CT145	Cardiovascular devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	29/08/2023	29/08/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	29/08/2023	29/08/2023
					CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	29/08/2023	29/08/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	29/08/2023	29/08/2023
					CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	29/08/2023	29/08/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	29/08/2023	29/08/2023
					CT233	Surgical	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	29/08/2023	29/08/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	29/08/2023	29/08/2023
					CT2406	Implantable	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	29/08/2023	29/08/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	29/08/2023	29/08/2023
					CT2407	Active implantable	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	29/08/2023	29/08/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	29/08/2023	29/08/2023
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	29/08/2023	29/08/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	29/08/2023	29/08/2023
					CT2727	Active implant devices/ancillaries	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	29/08/2023	29/08/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	29/08/2023	29/08/2023
					CT2750	Implantable cardiac electrical stimulators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	29/08/2023	29/08/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	29/08/2023	29/08/2023
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	29/08/2023	29/08/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	29/08/2023	29/08/2023
					CT2985	Active	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	29/08/2023	29/08/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	29/08/2023	29/08/2023
					CT3001	Implantable metal	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	29/08/2023	29/08/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	29/08/2023	29/08/2023
					CT3002	Implantable synthetic polymer	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	29/08/2023	29/08/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	29/08/2023	29/08/2023
					CT3006	Implantable inorganic material	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	29/08/2023	29/08/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	29/08/2023	29/08/2023
					CT314	Electrically powered	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	29/08/2023	29/08/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	29/08/2023	29/08/2023
					CT321	Long-term surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	29/08/2023	29/08/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	29/08/2023	29/08/2023
					CT449	Generators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	29/08/2023	29/08/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	29/08/2023	29/08/2023
					CT981	Single use	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	29/08/2023	29/08/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	29/08/2023	29/08/2023
					CT983	Surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	29/08/2023	29/08/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	29/08/2023	29/08/2023
14193	2023	All other reportable incidents	35223	Endocardial/interventricular septal pacing lead	CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070909	Oversensing		
							F26	No Health Consequences or Impact	A070909	Oversensing		
14194	2023	All other reportable incidents	47270	Cardiac resynchronization therapy implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E0119	Loss of Consciousness	A070909	Oversensing		
									A071201	Failure to Convert Rhythm		
									A0713	Defibrillation/Stimulation Problem		
									A072202	Low Impedance		
									A090801	Signal Artifact/Noise		
									A0909	Unexpected Therapeutic Results		
							E011903	Syncope/Fainting	A070909	Oversensing		
									A071201	Failure to Convert Rhythm		
									A0713	Defibrillation/Stimulation Problem		

									A072202	Low Impedance		
									A090801	Signal Artifact/Noise		
									A0909	Unexpected Therapeutic Results		
							E2104	Electric Shock	A070909	Oversensing		
									A071201	Failure to Convert Rhythm		
									A0713	Defibrillation/Stimulation Problem		
									A072202	Low Impedance		
									A090801	Signal Artifact/Noise		
									A0909	Unexpected Therapeutic Results		
							F05	Delay to Treatment/ Therapy	A070909	Oversensing		
									A071201	Failure to Convert Rhythm		
									A0713	Defibrillation/Stimulation Problem		
									A072202	Low Impedance		
									A090801	Signal Artifact/Noise		
									A0909	Unexpected Therapeutic Results		
							F08	Hospitalization or Prolonged Hospitalization	A070909	Oversensing		
									A071201	Failure to Convert Rhythm		
									A0713	Defibrillation/Stimulation Problem		
									A072202	Low Impedance		
									A090801	Signal Artifact/Noise		
									A0909	Unexpected Therapeutic Results		
							F1905	Device Revision or Replacement	A070909	Oversensing		
									A071201	Failure to Convert Rhythm		
									A0713	Defibrillation/Stimulation Problem		
									A072202	Low Impedance		
									A090801	Signal Artifact/Noise		
									A0909	Unexpected Therapeutic Results		
14195	2023	All other reportable incidents	47265	Dual-chamber implantable pacemaker, rate-responsive	CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
14196	2023	All other reportable incidents	47265	Dual-chamber implantable pacemaker, rate-responsive	CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	20/09/2023	20/09/2023
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	20/09/2023	20/09/2023
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	20/09/2023	20/09/2023
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	20/09/2023	20/09/2023
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	20/09/2023	20/09/2023
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	20/09/2023	20/09/2023
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	20/09/2023	20/09/2023
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	20/09/2023	20/09/2023
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	20/09/2023	20/09/2023
14197	2023	All other reportable incidents	47265	Dual-chamber implantable pacemaker, rate-responsive	CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem	28/09/2023	28/09/2023
									A070504	Premature Discharge of Battery	28/09/2023	28/09/2023
							F08	Hospitalization or Prolonged Hospitalization	A0705	Battery Problem	28/09/2023	28/09/2023
									A070504	Premature Discharge of Battery	28/09/2023	28/09/2023
							F1905	Device Revision or Replacement	A0705	Battery Problem	28/09/2023	28/09/2023
									A070504	Premature Discharge of Battery	28/09/2023	28/09/2023
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem	28/09/2023	28/09/2023
									A070504	Premature Discharge of Battery	28/09/2023	28/09/2023
							F08	Hospitalization or Prolonged Hospitalization	A0705	Battery Problem	28/09/2023	28/09/2023
									A070504	Premature Discharge of Battery	28/09/2023	28/09/2023
							F1905	Device Revision or Replacement	A0705	Battery Problem	28/09/2023	28/09/2023
									A070504	Premature Discharge of Battery	28/09/2023	28/09/2023
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem	28/09/2023	28/09/2023
									A070504	Premature Discharge of Battery	28/09/2023	28/09/2023
							F08	Hospitalization or Prolonged Hospitalization	A0705	Battery Problem	28/09/2023	28/09/2023
									A070504	Premature Discharge of Battery	28/09/2023	28/09/2023
							F1905	Device Revision or Replacement	A0705	Battery Problem	28/09/2023	28/09/2023
									A070504	Premature Discharge of Battery	28/09/2023	28/09/2023
14198	2023	All other reportable incidents	47265	Dual-chamber implantable pacemaker, rate-responsive	CT141	Electrodes/Leads and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A0908	Incorrect, Inadequate or Imprecise Result or Readings	06/10/2023	06/10/2023
							F08	Hospitalization or Prolonged Hospitalization	A0908	Incorrect, Inadequate or Imprecise Result or Readings	06/10/2023	06/10/2023
							F1905	Device Revision or Replacement	A0908	Incorrect, Inadequate or Imprecise Result or Readings	06/10/2023	06/10/2023
					CT145	Cardiovascular devices	E2403	No Clinical Signs, Symptoms or Conditions	A0908	Incorrect, Inadequate or Imprecise Result or Readings	06/10/2023	06/10/2023
							F08	Hospitalization or Prolonged Hospitalization	A0908	Incorrect, Inadequate or Imprecise Result or Readings	06/10/2023	06/10/2023
							F1905	Device Revision or Replacement	A0908	Incorrect, Inadequate or Imprecise Result or Readings	06/10/2023	06/10/2023
					CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A0908	Incorrect, Inadequate or Imprecise Result or Readings	06/10/2023	06/10/2023
							F08	Hospitalization or Prolonged Hospitalization	A0908	Incorrect, Inadequate or Imprecise Result or Readings	06/10/2023	06/10/2023
							F1905	Device Revision or Replacement	A0908	Incorrect, Inadequate or Imprecise Result or Readings	06/10/2023	06/10/2023
					CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A0908	Incorrect, Inadequate or Imprecise Result or Readings	06/10/2023	06/10/2023
							F08	Hospitalization or Prolonged Hospitalization	A0908	Incorrect, Inadequate or Imprecise Result or Readings	06/10/2023	06/10/2023
							F1905	Device Revision or Replacement	A0908	Incorrect, Inadequate or Imprecise Result or Readings	06/10/2023	06/10/2023

					CT233	Surgical	E2403	No Clinical Signs, Symptoms or Conditions	A0908	Incorrect, Inadequate or Imprecise Result or Readings	06/10/2023	06/10/2023
							F08	Hospitalization or Prolonged Hospitalization	A0908	Incorrect, Inadequate or Imprecise Result or Readings	06/10/2023	06/10/2023
							F1905	Device Revision or Replacement	A0908	Incorrect, Inadequate or Imprecise Result or Readings	06/10/2023	06/10/2023
					CT2406	Implantable	E2403	No Clinical Signs, Symptoms or Conditions	A0908	Incorrect, Inadequate or Imprecise Result or Readings	06/10/2023	06/10/2023
							F08	Hospitalization or Prolonged Hospitalization	A0908	Incorrect, Inadequate or Imprecise Result or Readings	06/10/2023	06/10/2023
							F1905	Device Revision or Replacement	A0908	Incorrect, Inadequate or Imprecise Result or Readings	06/10/2023	06/10/2023
					CT2407	Active implantable	E2403	No Clinical Signs, Symptoms or Conditions	A0908	Incorrect, Inadequate or Imprecise Result or Readings	06/10/2023	06/10/2023
							F08	Hospitalization or Prolonged Hospitalization	A0908	Incorrect, Inadequate or Imprecise Result or Readings	06/10/2023	06/10/2023
							F1905	Device Revision or Replacement	A0908	Incorrect, Inadequate or Imprecise Result or Readings	06/10/2023	06/10/2023
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A0908	Incorrect, Inadequate or Imprecise Result or Readings	06/10/2023	06/10/2023
							F08	Hospitalization or Prolonged Hospitalization	A0908	Incorrect, Inadequate or Imprecise Result or Readings	06/10/2023	06/10/2023
							F1905	Device Revision or Replacement	A0908	Incorrect, Inadequate or Imprecise Result or Readings	06/10/2023	06/10/2023
					CT2727	Active implant devices/ancillaries	E2403	No Clinical Signs, Symptoms or Conditions	A0908	Incorrect, Inadequate or Imprecise Result or Readings	06/10/2023	06/10/2023
							F08	Hospitalization or Prolonged Hospitalization	A0908	Incorrect, Inadequate or Imprecise Result or Readings	06/10/2023	06/10/2023
							F1905	Device Revision or Replacement	A0908	Incorrect, Inadequate or Imprecise Result or Readings	06/10/2023	06/10/2023
					CT2750	Implantable cardiac electrical stimulators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A0908	Incorrect, Inadequate or Imprecise Result or Readings	06/10/2023	06/10/2023
							F08	Hospitalization or Prolonged Hospitalization	A0908	Incorrect, Inadequate or Imprecise Result or Readings	06/10/2023	06/10/2023
							F1905	Device Revision or Replacement	A0908	Incorrect, Inadequate or Imprecise Result or Readings	06/10/2023	06/10/2023
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A0908	Incorrect, Inadequate or Imprecise Result or Readings	06/10/2023	06/10/2023
							F08	Hospitalization or Prolonged Hospitalization	A0908	Incorrect, Inadequate or Imprecise Result or Readings	06/10/2023	06/10/2023
							F1905	Device Revision or Replacement	A0908	Incorrect, Inadequate or Imprecise Result or Readings	06/10/2023	06/10/2023
					CT2985	Active	E2403	No Clinical Signs, Symptoms or Conditions	A0908	Incorrect, Inadequate or Imprecise Result or Readings	06/10/2023	06/10/2023
							F08	Hospitalization or Prolonged Hospitalization	A0908	Incorrect, Inadequate or Imprecise Result or Readings	06/10/2023	06/10/2023
							F1905	Device Revision or Replacement	A0908	Incorrect, Inadequate or Imprecise Result or Readings	06/10/2023	06/10/2023
					CT3001	Implantable metal	E2403	No Clinical Signs, Symptoms or Conditions	A0908	Incorrect, Inadequate or Imprecise Result or Readings	06/10/2023	06/10/2023
							F08	Hospitalization or Prolonged Hospitalization	A0908	Incorrect, Inadequate or Imprecise Result or Readings	06/10/2023	06/10/2023
							F1905	Device Revision or Replacement	A0908	Incorrect, Inadequate or Imprecise Result or Readings	06/10/2023	06/10/2023
					CT3002	Implantable synthetic polymer	E2403	No Clinical Signs, Symptoms or Conditions	A0908	Incorrect, Inadequate or Imprecise Result or Readings	06/10/2023	06/10/2023
							F08	Hospitalization or Prolonged Hospitalization	A0908	Incorrect, Inadequate or Imprecise Result or Readings	06/10/2023	06/10/2023
							F1905	Device Revision or Replacement	A0908	Incorrect, Inadequate or Imprecise Result or Readings	06/10/2023	06/10/2023
					CT3006	Implantable inorganic material	E2403	No Clinical Signs, Symptoms or Conditions	A0908	Incorrect, Inadequate or Imprecise Result or Readings	06/10/2023	06/10/2023
							F08	Hospitalization or Prolonged Hospitalization	A0908	Incorrect, Inadequate or Imprecise Result or Readings	06/10/2023	06/10/2023
							F1905	Device Revision or Replacement	A0908	Incorrect, Inadequate or Imprecise Result or Readings	06/10/2023	06/10/2023
					CT314	Electrically powered	E2403	No Clinical Signs, Symptoms or Conditions	A0908	Incorrect, Inadequate or Imprecise Result or Readings	06/10/2023	06/10/2023
							F08	Hospitalization or Prolonged Hospitalization	A0908	Incorrect, Inadequate or Imprecise Result or Readings	06/10/2023	06/10/2023
							F1905	Device Revision or Replacement	A0908	Incorrect, Inadequate or Imprecise Result or Readings	06/10/2023	06/10/2023
					CT321	Long-term surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A0908	Incorrect, Inadequate or Imprecise Result or Readings	06/10/2023	06/10/2023
							F08	Hospitalization or Prolonged Hospitalization	A0908	Incorrect, Inadequate or Imprecise Result or Readings	06/10/2023	06/10/2023
							F1905	Device Revision or Replacement	A0908	Incorrect, Inadequate or Imprecise Result or Readings	06/10/2023	06/10/2023
					CT449	Generators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A0908	Incorrect, Inadequate or Imprecise Result or Readings	06/10/2023	06/10/2023
							F08	Hospitalization or Prolonged Hospitalization	A0908	Incorrect, Inadequate or Imprecise Result or Readings	06/10/2023	06/10/2023
							F1905	Device Revision or Replacement	A0908	Incorrect, Inadequate or Imprecise Result or Readings	06/10/2023	06/10/2023
					CT981	Single use	E2403	No Clinical Signs, Symptoms or Conditions	A0908	Incorrect, Inadequate or Imprecise Result or Readings	06/10/2023	06/10/2023
							F08	Hospitalization or Prolonged Hospitalization	A0908	Incorrect, Inadequate or Imprecise Result or Readings	06/10/2023	06/10/2023
							F1905	Device Revision or Replacement	A0908	Incorrect, Inadequate or Imprecise Result or Readings	06/10/2023	06/10/2023
					CT983	Surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A0908	Incorrect, Inadequate or Imprecise Result or Readings	06/10/2023	06/10/2023
							F08	Hospitalization or Prolonged Hospitalization	A0908	Incorrect, Inadequate or Imprecise Result or Readings	06/10/2023	06/10/2023
							F1905	Device Revision or Replacement	A0908	Incorrect, Inadequate or Imprecise Result or Readings	06/10/2023	06/10/2023
14199	2023	All other reportable incidents	35852	Single-chamber implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	15/02/2024	15/02/2024
									A0908	Incorrect, Inadequate or Imprecise Result or Readings	15/02/2024	15/02/2024

							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	15/02/2024	15/02/2024
									A0908	Incorrect, Inadequate or Imprecise Result or Readings	15/02/2024	15/02/2024
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	15/02/2024	15/02/2024
									A0908	Incorrect, Inadequate or Imprecise Result or Readings	15/02/2024	15/02/2024
14200	2023	All other reportable incidents	35852	Single-chamber implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E2104	Electric Shock	A070909	Oversensing		
									A071302	Inappropriate/Inadequate Shock/Stimulation		
									A1301	Failure to Read Input Signal		
									A23	Use of Device Problem		
							F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A070909	Oversensing		
									A071302	Inappropriate/Inadequate Shock/Stimulation		
									A1301	Failure to Read Input Signal		
									A23	Use of Device Problem		
14201	2023	All other reportable incidents	47265	Dual-chamber implantable pacemaker, rate-responsive	CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
14202	2023		47265	Dual-chamber implantable pacemaker, rate-responsive	CT141	Electrodes/Leads and associated devices	E0601	Arrhythmia	A070504	Premature Discharge of Battery	29/09/2023	29/09/2023
									A070909	Oversensing	29/09/2023	29/09/2023
									A071102	Failure to Interrogate	29/09/2023	29/09/2023
									A071204	No Pacing	29/09/2023	29/09/2023
							F1901	Additional Surgery	A070504	Premature Discharge of Battery	29/09/2023	29/09/2023
									A070909	Oversensing	29/09/2023	29/09/2023
									A071102	Failure to Interrogate	29/09/2023	29/09/2023
									A071204	No Pacing	29/09/2023	29/09/2023
					CT145	Cardiovascular devices	E0601	Arrhythmia	A070504	Premature Discharge of Battery	29/09/2023	29/09/2023
									A070909	Oversensing	29/09/2023	29/09/2023
									A071102	Failure to Interrogate	29/09/2023	29/09/2023
									A071204	No Pacing	29/09/2023	29/09/2023
							F1901	Additional Surgery	A070504	Premature Discharge of Battery	29/09/2023	29/09/2023
									A070909	Oversensing	29/09/2023	29/09/2023
									A071102	Failure to Interrogate	29/09/2023	29/09/2023
									A071204	No Pacing	29/09/2023	29/09/2023
					CT1684	Cardiac pacemakers and associated devices	E0601	Arrhythmia	A070504	Premature Discharge of Battery	29/09/2023	29/09/2023
									A070909	Oversensing	29/09/2023	29/09/2023
									A071102	Failure to Interrogate	29/09/2023	29/09/2023
									A071204	No Pacing	29/09/2023	29/09/2023
							F1901	Additional Surgery	A070504	Premature Discharge of Battery	29/09/2023	29/09/2023
									A070909	Oversensing	29/09/2023	29/09/2023
									A071102	Failure to Interrogate	29/09/2023	29/09/2023
									A071204	No Pacing	29/09/2023	29/09/2023
					CT1685	Cardiac pacemakers	E0601	Arrhythmia	A070504	Premature Discharge of Battery	29/09/2023	29/09/2023
									A070909	Oversensing	29/09/2023	29/09/2023
									A071102	Failure to Interrogate	29/09/2023	29/09/2023
									A071204	No Pacing	29/09/2023	29/09/2023
							F1901	Additional Surgery	A070504	Premature Discharge of Battery	29/09/2023	29/09/2023
									A070909	Oversensing	29/09/2023	29/09/2023
									A071102	Failure to Interrogate	29/09/2023	29/09/2023
									A071204	No Pacing	29/09/2023	29/09/2023
					CT233	Surgical	E0601	Arrhythmia	A070504	Premature Discharge of Battery	29/09/2023	29/09/2023
									A070909	Oversensing	29/09/2023	29/09/2023
									A071102	Failure to Interrogate	29/09/2023	29/09/2023
									A071204	No Pacing	29/09/2023	29/09/2023
							F1901	Additional Surgery	A070504	Premature Discharge of Battery	29/09/2023	29/09/2023
									A070909	Oversensing	29/09/2023	29/09/2023
									A071102	Failure to Interrogate	29/09/2023	29/09/2023
									A071204	No Pacing	29/09/2023	29/09/2023
					CT2406	Implantable	E0601	Arrhythmia	A070504	Premature Discharge of Battery	29/09/2023	29/09/2023
									A070909	Oversensing	29/09/2023	29/09/2023
									A071102	Failure to Interrogate	29/09/2023	29/09/2023
									A071204	No Pacing	29/09/2023	29/09/2023
							F1901	Additional Surgery	A070504	Premature Discharge of Battery	29/09/2023	29/09/2023
									A070909	Oversensing	29/09/2023	29/09/2023
									A071102	Failure to Interrogate	29/09/2023	29/09/2023
									A071204	No Pacing	29/09/2023	29/09/2023
					CT2407	Active implantable	E0601	Arrhythmia	A070504	Premature Discharge of Battery	29/09/2023	29/09/2023
									A070909	Oversensing	29/09/2023	29/09/2023
									A071102	Failure to Interrogate	29/09/2023	29/09/2023
									A071204	No Pacing	29/09/2023	29/09/2023
							F1901	Additional Surgery	A070504	Premature Discharge of Battery	29/09/2023	29/09/2023
									A070909	Oversensing	29/09/2023	29/09/2023
									A071102	Failure to Interrogate	29/09/2023	29/09/2023
									A071204	No Pacing	29/09/2023	29/09/2023
					CT257	Implantable cardiac pacemakers	E0601	Arrhythmia	A070504	Premature Discharge of Battery	29/09/2023	29/09/2023
									A070909	Oversensing	29/09/2023	29/09/2023
									A071102	Failure to Interrogate	29/09/2023	29/09/2023
									A071204	No Pacing	29/09/2023	29/09/2023
							F1901	Additional Surgery	A070504	Premature Discharge of Battery	29/09/2023	29/09/2023
									A070909	Oversensing	29/09/2023	29/09/2023

								A071102	Failure to Interrogate	29/09/2023	29/09/2023
								A071204	No Pacing	29/09/2023	29/09/2023
					CT2727	Active implant devices/ancillaries	E0601	Arrhythmia	A070504	Premature Discharge of Battery	29/09/2023	29/09/2023
									A070909	Oversensing	29/09/2023	29/09/2023
									A071102	Failure to Interrogate	29/09/2023	29/09/2023
									A071204	No Pacing	29/09/2023	29/09/2023
							F1901	Additional Surgery	A070504	Premature Discharge of Battery	29/09/2023	29/09/2023
									A070909	Oversensing	29/09/2023	29/09/2023
									A071102	Failure to Interrogate	29/09/2023	29/09/2023
									A071204	No Pacing	29/09/2023	29/09/2023
					CT2750	Implantable cardiac electrical stimulators and associated devices	E0601	Arrhythmia	A070504	Premature Discharge of Battery	29/09/2023	29/09/2023
									A070909	Oversensing	29/09/2023	29/09/2023
									A071102	Failure to Interrogate	29/09/2023	29/09/2023
									A071204	No Pacing	29/09/2023	29/09/2023
							F1901	Additional Surgery	A070504	Premature Discharge of Battery	29/09/2023	29/09/2023
									A070909	Oversensing	29/09/2023	29/09/2023
									A071102	Failure to Interrogate	29/09/2023	29/09/2023
									A071204	No Pacing	29/09/2023	29/09/2023
					CT2866	Implantable cardiac electrical stimulators	E0601	Arrhythmia	A070504	Premature Discharge of Battery	29/09/2023	29/09/2023
									A070909	Oversensing	29/09/2023	29/09/2023
									A071102	Failure to Interrogate	29/09/2023	29/09/2023
									A071204	No Pacing	29/09/2023	29/09/2023
							F1901	Additional Surgery	A070504	Premature Discharge of Battery	29/09/2023	29/09/2023
									A070909	Oversensing	29/09/2023	29/09/2023
									A071102	Failure to Interrogate	29/09/2023	29/09/2023
									A071204	No Pacing	29/09/2023	29/09/2023
					CT2985	Active	E0601	Arrhythmia	A070504	Premature Discharge of Battery	29/09/2023	29/09/2023
									A070909	Oversensing	29/09/2023	29/09/2023
									A071102	Failure to Interrogate	29/09/2023	29/09/2023
									A071204	No Pacing	29/09/2023	29/09/2023
							F1901	Additional Surgery	A070504	Premature Discharge of Battery	29/09/2023	29/09/2023
									A070909	Oversensing	29/09/2023	29/09/2023
									A071102	Failure to Interrogate	29/09/2023	29/09/2023
									A071204	No Pacing	29/09/2023	29/09/2023
					CT3001	Implantable metal	E0601	Arrhythmia	A070504	Premature Discharge of Battery	29/09/2023	29/09/2023
									A070909	Oversensing	29/09/2023	29/09/2023
									A071102	Failure to Interrogate	29/09/2023	29/09/2023
									A071204	No Pacing	29/09/2023	29/09/2023
							F1901	Additional Surgery	A070504	Premature Discharge of Battery	29/09/2023	29/09/2023
									A070909	Oversensing	29/09/2023	29/09/2023
									A071102	Failure to Interrogate	29/09/2023	29/09/2023
									A071204	No Pacing	29/09/2023	29/09/2023
					CT3002	Implantable synthetic polymer	E0601	Arrhythmia	A070504	Premature Discharge of Battery	29/09/2023	29/09/2023
									A070909	Oversensing	29/09/2023	29/09/2023
									A071102	Failure to Interrogate	29/09/2023	29/09/2023
									A071204	No Pacing	29/09/2023	29/09/2023
							F1901	Additional Surgery	A070504	Premature Discharge of Battery	29/09/2023	29/09/2023
									A070909	Oversensing	29/09/2023	29/09/2023
									A071102	Failure to Interrogate	29/09/2023	29/09/2023
									A071204	No Pacing	29/09/2023	29/09/2023
					CT3006	Implantable inorganic material	E0601	Arrhythmia	A070504	Premature Discharge of Battery	29/09/2023	29/09/2023
									A070909	Oversensing	29/09/2023	29/09/2023
									A071102	Failure to Interrogate	29/09/2023	29/09/2023
									A071204	No Pacing	29/09/2023	29/09/2023
							F1901	Additional Surgery	A070504	Premature Discharge of Battery	29/09/2023	29/09/2023
									A070909	Oversensing	29/09/2023	29/09/2023
									A071102	Failure to Interrogate	29/09/2023	29/09/2023
									A071204	No Pacing	29/09/2023	29/09/2023
					CT314	Electrically powered	E0601	Arrhythmia	A070504	Premature Discharge of Battery	29/09/2023	29/09/2023
									A070909	Oversensing	29/09/2023	29/09/2023
									A071102	Failure to Interrogate	29/09/2023	29/09/2023
									A071204	No Pacing	29/09/2023	29/09/2023
							F1901	Additional Surgery	A070504	Premature Discharge of Battery	29/09/2023	29/09/2023
									A070909	Oversensing	29/09/2023	29/09/2023
									A071102	Failure to Interrogate	29/09/2023	29/09/2023
									A071204	No Pacing	29/09/2023	29/09/2023
					CT321	Long-term surgical invasive	E0601	Arrhythmia	A070504	Premature Discharge of Battery	29/09/2023	29/09/2023
									A070909	Oversensing	29/09/2023	29/09/2023
									A071102	Failure to Interrogate	29/09/2023	29/09/2023
									A071204	No Pacing	29/09/2023	29/09/2023
							F1901	Additional Surgery	A070504	Premature Discharge of Battery	29/09/2023	29/09/2023
									A070909	Oversensing	29/09/2023	29/09/2023
									A071102	Failure to Interrogate	29/09/2023	29/09/2023
									A071204	No Pacing	29/09/2023	29/09/2023
					CT449	Generators and associated devices	E0601	Arrhythmia	A070504	Premature Discharge of Battery	29/09/2023	29/09/2023
									A070909	Oversensing	29/09/2023	29/09/2023
									A071102	Failure to Interrogate	29/09/2023	29/09/2023
									A071204	No Pacing	29/09/2023	29/09/2023
							F1901	Additional Surgery	A070504	Premature Discharge of Battery	29/09/2023	29/09/2023
									A070909	Oversensing	29/09/2023	29/09/2023
									A071102	Failure to Interrogate	29/09/2023	29/09/2023
									A071204	No Pacing	29/09/2023	29/09/2023
					CT981	Single use	E0601	Arrhythmia	A070504	Premature Discharge of Battery	29/09/2023	29/09/2023
									A070909	Oversensing	29/09/2023	29/09/2023
									A071102	Failure to Interrogate	29/09/2023	29/09/2023
									A071204	No Pacing	29/09/2023	29/09/2023
							F1901	Additional Surgery	A070504	Premature Discharge of Battery	29/09/2023	29/09/2023
									A070909	Oversensing	29/09/2023	29/09/2023

									A071102	Failure to Interrogate	29/09/2023	29/09/2023
									A071204	No Pacing	29/09/2023	29/09/2023
					CT983	Surgical invasive	E0601	Arrhythmia	A070504	Premature Discharge of Battery	29/09/2023	29/09/2023
									A070909	Oversensing	29/09/2023	29/09/2023
									A071102	Failure to Interrogate	29/09/2023	29/09/2023
									A071204	No Pacing	29/09/2023	29/09/2023
							F1901	Additional Surgery	A070504	Premature Discharge of Battery	29/09/2023	29/09/2023
									A070909	Oversensing	29/09/2023	29/09/2023
									A071102	Failure to Interrogate	29/09/2023	29/09/2023
									A071204	No Pacing	29/09/2023	29/09/2023
14203	2023	All other reportable incidents	17882	Physiologic-monitoring defibrillation system	CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A07	Electrical /Electronic Property Problem		
							F27	Problem Identified During Non-Clinical Procedure	A07	Electrical /Electronic Property Problem		
14204	2023	All other reportable incidents	47265	Dual-chamber implantable pacemaker, rate-responsive	CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	04/06/2023	04/06/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	04/06/2023	04/06/2023
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	04/06/2023	04/06/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	04/06/2023	04/06/2023
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	04/06/2023	04/06/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	04/06/2023	04/06/2023
14205	2023	All other reportable incidents	35223	Endocardial/interventricular septal pacing lead	CT1684	Cardiac pacemakers and associated devices	E2311	Discomfort	A070909	Oversensing		
									A0712	Pacing Problem		
							F1901	Additional Surgery	A070909	Oversensing		
									A0712	Pacing Problem		
14206	2023	All other reportable incidents	35223	Endocardial/interventricular septal pacing lead	CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A07	Electrical /Electronic Property Problem		
									A072202	Low Impedance		
							F26	No Health Consequences or Impact	A07	Electrical /Electronic Property Problem		
									A072202	Low Impedance		
14207	2023	All other reportable incidents	47265	Dual-chamber implantable pacemaker, rate-responsive	CT141	Electrodes/Leads and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	01/08/2023	31/08/2023
							F1901	Additional Surgery	A070504	Premature Discharge of Battery	01/08/2023	31/08/2023
					CT145	Cardiovascular devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	01/08/2023	31/08/2023
							F1901	Additional Surgery	A070504	Premature Discharge of Battery	01/08/2023	31/08/2023
					CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	01/08/2023	31/08/2023
							F1901	Additional Surgery	A070504	Premature Discharge of Battery	01/08/2023	31/08/2023
					CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	01/08/2023	31/08/2023
							F1901	Additional Surgery	A070504	Premature Discharge of Battery	01/08/2023	31/08/2023
					CT233	Surgical	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	01/08/2023	31/08/2023
							F1901	Additional Surgery	A070504	Premature Discharge of Battery	01/08/2023	31/08/2023
					CT2406	Implantable	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	01/08/2023	31/08/2023
							F1901	Additional Surgery	A070504	Premature Discharge of Battery	01/08/2023	31/08/2023
					CT2407	Active implantable	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	01/08/2023	31/08/2023
							F1901	Additional Surgery	A070504	Premature Discharge of Battery	01/08/2023	31/08/2023
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	01/08/2023	31/08/2023
							F1901	Additional Surgery	A070504	Premature Discharge of Battery	01/08/2023	31/08/2023
					CT2727	Active implant devices/ancillaries	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	01/08/2023	31/08/2023
							F1901	Additional Surgery	A070504	Premature Discharge of Battery	01/08/2023	31/08/2023
					CT2750	Implantable cardiac electrical stimulators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	01/08/2023	31/08/2023
							F1901	Additional Surgery	A070504	Premature Discharge of Battery	01/08/2023	31/08/2023
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	01/08/2023	31/08/2023
							F1901	Additional Surgery	A070504	Premature Discharge of Battery	01/08/2023	31/08/2023
					CT2985	Active	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	01/08/2023	31/08/2023
							F1901	Additional Surgery	A070504	Premature Discharge of Battery	01/08/2023	31/08/2023
					CT3001	Implantable metal	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	01/08/2023	31/08/2023
							F1901	Additional Surgery	A070504	Premature Discharge of Battery	01/08/2023	31/08/2023
					CT3002	Implantable synthetic polymer	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	01/08/2023	31/08/2023
							F1901	Additional Surgery	A070504	Premature Discharge of Battery	01/08/2023	31/08/2023
					CT3006	Implantable inorganic material	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	01/08/2023	31/08/2023
							F1901	Additional Surgery	A070504	Premature Discharge of Battery	01/08/2023	31/08/2023
					CT314	Electrically powered	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	01/08/2023	31/08/2023
							F1901	Additional Surgery	A070504	Premature Discharge of Battery	01/08/2023	31/08/2023
					CT321	Long-term surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	01/08/2023	31/08/2023
							F1901	Additional Surgery	A070504	Premature Discharge of Battery	01/08/2023	31/08/2023
					CT449	Generators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	01/08/2023	31/08/2023
							F1901	Additional Surgery	A070504	Premature Discharge of Battery	01/08/2023	31/08/2023
					CT981	Single use	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	01/08/2023	31/08/2023
							F1901	Additional Surgery	A070504	Premature Discharge of Battery	01/08/2023	31/08/2023
					CT983	Surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	01/08/2023	31/08/2023
							F1901	Additional Surgery	A070504	Premature Discharge of Battery	01/08/2023	31/08/2023
14208	2023	All other reportable incidents	47265	Dual-chamber implantable pacemaker, rate-responsive	CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A071102	Failure to Interrogate	13/07/2023	13/07/2023
							F1901	Additional Surgery	A071102	Failure to Interrogate	13/07/2023	13/07/2023
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A071102	Failure to Interrogate	13/07/2023	13/07/2023
							F1901	Additional Surgery	A071102	Failure to Interrogate	13/07/2023	13/07/2023
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A071102	Failure to Interrogate	13/07/2023	13/07/2023
							F1901	Additional Surgery	A071102	Failure to Interrogate	13/07/2023	13/07/2023
14209	2023	All other reportable incidents	17882	Physiologic-monitoring defibrillation system	CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A0719	Unexpected Shutdown		
							F27	Problem Identified During Non-Clinical Procedure	A0719	Unexpected Shutdown		
14210	2023		47265	Dual-chamber implantable pacemaker, rate-responsive	CT141	Electrodes/Leads and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	10/10/2023	10/10/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	10/10/2023	10/10/2023
					CT145	Cardiovascular devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	10/10/2023	10/10/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	10/10/2023	10/10/2023
					CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	10/10/2023	10/10/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	10/10/2023	10/10/2023
					CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	10/10/2023	10/10/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	10/10/2023	10/10/2023
					CT233	Surgical	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	10/10/2023	10/10/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	10/10/2023	10/10/2023
					CT2406	Implantable	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	10/10/2023	10/10/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	10/10/2023	10/10/2023
					CT2407	Active implantable	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	10/10/2023	10/10/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	10/10/2023	10/10/2023

					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	10/10/2023	10/10/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	10/10/2023	10/10/2023
					CT2727	Active implant devices/ancillaries	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	10/10/2023	10/10/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	10/10/2023	10/10/2023
					CT2750	Implantable cardiac electrical stimulators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	10/10/2023	10/10/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	10/10/2023	10/10/2023
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	10/10/2023	10/10/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	10/10/2023	10/10/2023
					CT2985	Active	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	10/10/2023	10/10/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	10/10/2023	10/10/2023
					CT3001	Implantable metal	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	10/10/2023	10/10/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	10/10/2023	10/10/2023
					CT3002	Implantable synthetic polymer	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	10/10/2023	10/10/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	10/10/2023	10/10/2023
					CT3006	Implantable inorganic material	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	10/10/2023	10/10/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	10/10/2023	10/10/2023
					CT314	Electrically powered	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	10/10/2023	10/10/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	10/10/2023	10/10/2023
					CT321	Long-term surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	10/10/2023	10/10/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	10/10/2023	10/10/2023
					CT449	Generators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	10/10/2023	10/10/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	10/10/2023	10/10/2023
					CT981	Single use	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	10/10/2023	10/10/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	10/10/2023	10/10/2023
					CT983	Surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	10/10/2023	10/10/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	10/10/2023	10/10/2023
14211	2023	All other reportable incidents	17882	Physiologic-monitoring defibrillation system	CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem		
							F27	Problem Identified During Non-Clinical Procedure	A0705	Battery Problem		
14212	2023	All other reportable incidents	47265	Dual-chamber implantable pacemaker, rate-responsive	CT1685	Cardiac pacemakers	E060104	Bradycardia	A070101	Failure to Capture	21/08/2023	21/08/2023
							F1905	Device Revision or Replacement	A070101	Failure to Capture	21/08/2023	21/08/2023
					CT257	Implantable cardiac pacemakers	E060104	Bradycardia	A070101	Failure to Capture	21/08/2023	21/08/2023
							F1905	Device Revision or Replacement	A070101	Failure to Capture	21/08/2023	21/08/2023
					CT2866	Implantable cardiac electrical stimulators	E060104	Bradycardia	A070101	Failure to Capture	21/08/2023	21/08/2023
							F1905	Device Revision or Replacement	A070101	Failure to Capture	21/08/2023	21/08/2023
14213	2023		47265	Dual-chamber implantable pacemaker, rate-responsive	CT141	Electrodes/Leads and associated devices	E011902	Presyncope	A07	Electrical /Electronic Property Problem		
									A0705	Battery Problem		
							E012207	Twitching	A07	Electrical /Electronic Property Problem		
									A0705	Battery Problem		
							E06	Heart	A07	Electrical /Electronic Property Problem		
									A0705	Battery Problem		
							F01	Change in Therapeutic Response	A07	Electrical /Electronic Property Problem		
									A0705	Battery Problem		
							F19	Surgical Intervention	A07	Electrical /Electronic Property Problem		
									A0705	Battery Problem		
							F1905	Device Revision or Replacement	A07	Electrical /Electronic Property Problem		
									A0705	Battery Problem		
							F23	Unexpected Medical Intervention	A07	Electrical /Electronic Property Problem		
									A0705	Battery Problem		
					CT145	Cardiovascular devices	E011902	Presyncope	A07	Electrical /Electronic Property Problem		
									A0705	Battery Problem		
							E012207	Twitching	A07	Electrical /Electronic Property Problem		
									A0705	Battery Problem		
							E06	Heart	A07	Electrical /Electronic Property Problem		
									A0705	Battery Problem		
							F01	Change in Therapeutic Response	A07	Electrical /Electronic Property Problem		
									A0705	Battery Problem		
							F19	Surgical Intervention	A07	Electrical /Electronic Property Problem		
									A0705	Battery Problem		
							F1905	Device Revision or Replacement	A07	Electrical /Electronic Property Problem		
									A0705	Battery Problem		
							F23	Unexpected Medical Intervention	A07	Electrical /Electronic Property Problem		
									A0705	Battery Problem		
					CT1684	Cardiac pacemakers and associated devices	E011902	Presyncope	A07	Electrical /Electronic Property Problem		
									A0705	Battery Problem		
							E012207	Twitching	A07	Electrical /Electronic Property Problem		
									A0705	Battery Problem		
							E06	Heart	A07	Electrical /Electronic Property Problem		
									A0705	Battery Problem		
							F01	Change in Therapeutic Response	A07	Electrical /Electronic Property Problem		
									A0705	Battery Problem		
							F19	Surgical Intervention	A07	Electrical /Electronic Property Problem		
									A0705	Battery Problem		
							F1905	Device Revision or Replacement	A07	Electrical /Electronic Property Problem		
									A0705	Battery Problem		
							F23	Unexpected Medical Intervention	A07	Electrical /Electronic Property Problem		
									A0705	Battery Problem		
					CT1685	Cardiac pacemakers	E011902	Presyncope	A07	Electrical /Electronic Property Problem		
									A0705	Battery Problem		
							E012207	Twitching	A07	Electrical /Electronic Property Problem		
									A0705	Battery Problem		
							E06	Heart	A07	Electrical /Electronic Property Problem		
									A0705	Battery Problem		
							F01	Change in Therapeutic Response	A07	Electrical /Electronic Property Problem		
									A0705	Battery Problem		
							F19	Surgical Intervention	A07	Electrical /Electronic Property Problem		
									A0705	Battery Problem		
							F1905	Device Revision or Replacement	A07	Electrical /Electronic Property Problem		
									A0705	Battery Problem		

							F23	Unexpected Medical Intervention	A07	Electrical /Electronic Property Problem		
									A0705	Battery Problem		
					CT233	Surgical	E011902	Presyncope	A07	Electrical /Electronic Property Problem		
									A0705	Battery Problem		
							E012207	Twitching	A07	Electrical /Electronic Property Problem		
									A0705	Battery Problem		
							E06	Heart	A07	Electrical /Electronic Property Problem		
									A0705	Battery Problem		
							F01	Change in Therapeutic Response	A07	Electrical /Electronic Property Problem		
									A0705	Battery Problem		
							F19	Surgical Intervention	A07	Electrical /Electronic Property Problem		
									A0705	Battery Problem		
							F1905	Device Revision or Replacement	A07	Electrical /Electronic Property Problem		
									A0705	Battery Problem		
							F23	Unexpected Medical Intervention	A07	Electrical /Electronic Property Problem		
									A0705	Battery Problem		
					CT2406	Implantable	E011902	Presyncope	A07	Electrical /Electronic Property Problem		
									A0705	Battery Problem		
							E012207	Twitching	A07	Electrical /Electronic Property Problem		
									A0705	Battery Problem		
							E06	Heart	A07	Electrical /Electronic Property Problem		
									A0705	Battery Problem		
							F01	Change in Therapeutic Response	A07	Electrical /Electronic Property Problem		
									A0705	Battery Problem		
							F19	Surgical Intervention	A07	Electrical /Electronic Property Problem		
									A0705	Battery Problem		
							F1905	Device Revision or Replacement	A07	Electrical /Electronic Property Problem		
									A0705	Battery Problem		
							F23	Unexpected Medical Intervention	A07	Electrical /Electronic Property Problem		
									A0705	Battery Problem		
					CT2407	Active implantable	E011902	Presyncope	A07	Electrical /Electronic Property Problem		
									A0705	Battery Problem		
							E012207	Twitching	A07	Electrical /Electronic Property Problem		
									A0705	Battery Problem		
							E06	Heart	A07	Electrical /Electronic Property Problem		
									A0705	Battery Problem		
							F01	Change in Therapeutic Response	A07	Electrical /Electronic Property Problem		
									A0705	Battery Problem		
							F19	Surgical Intervention	A07	Electrical /Electronic Property Problem		
									A0705	Battery Problem		
							F1905	Device Revision or Replacement	A07	Electrical /Electronic Property Problem		
									A0705	Battery Problem		
							F23	Unexpected Medical Intervention	A07	Electrical /Electronic Property Problem		
									A0705	Battery Problem		
					CT257	Implantable cardiac pacemakers	E011902	Presyncope	A07	Electrical /Electronic Property Problem		
									A0705	Battery Problem		
							E012207	Twitching	A07	Electrical /Electronic Property Problem		
									A0705	Battery Problem		
							E06	Heart	A07	Electrical /Electronic Property Problem		
									A0705	Battery Problem		
							F01	Change in Therapeutic Response	A07	Electrical /Electronic Property Problem		
									A0705	Battery Problem		
							F19	Surgical Intervention	A07	Electrical /Electronic Property Problem		
									A0705	Battery Problem		
							F1905	Device Revision or Replacement	A07	Electrical /Electronic Property Problem		
									A0705	Battery Problem		
							F23	Unexpected Medical Intervention	A07	Electrical /Electronic Property Problem		
									A0705	Battery Problem		
					CT2727	Active implant devices/ancillaries	E011902	Presyncope	A07	Electrical /Electronic Property Problem		
									A0705	Battery Problem		
							E012207	Twitching	A07	Electrical /Electronic Property Problem		
									A0705	Battery Problem		
							E06	Heart	A07	Electrical /Electronic Property Problem		
									A0705	Battery Problem		
							F01	Change in Therapeutic Response	A07	Electrical /Electronic Property Problem		
									A0705	Battery Problem		
							F19	Surgical Intervention	A07	Electrical /Electronic Property Problem		
									A0705	Battery Problem		
							F1905	Device Revision or Replacement	A07	Electrical /Electronic Property Problem		
									A0705	Battery Problem		
							F23	Unexpected Medical Intervention	A07	Electrical /Electronic Property Problem		
									A0705	Battery Problem		
					CT2750	Implantable cardiac electrical stimulators and associated devices	E011902	Presyncope	A07	Electrical /Electronic Property Problem		
									A0705	Battery Problem		
							E012207	Twitching	A07	Electrical /Electronic Property Problem		
									A0705	Battery Problem		
							E06	Heart	A07	Electrical /Electronic Property Problem		
									A0705	Battery Problem		
							F01	Change in Therapeutic Response	A07	Electrical /Electronic Property Problem		
									A0705	Battery Problem		
							F19	Surgical Intervention	A07	Electrical /Electronic Property Problem		
									A0705	Battery Problem		
							F1905	Device Revision or Replacement	A07	Electrical /Electronic Property Problem		
									A0705	Battery Problem		
							F23	Unexpected Medical Intervention	A07	Electrical /Electronic Property Problem		
									A0705	Battery Problem		
					CT2866	Implantable cardiac electrical stimulators	E011902	Presyncope	A07	Electrical /Electronic Property Problem		
									A0705	Battery Problem		

							E012207	Twitching	A07	Electrical /Electronic Property Problem		
									A0705	Battery Problem		
							E06	Heart	A07	Electrical /Electronic Property Problem		
									A0705	Battery Problem		
							F01	Change in Therapeutic Response	A07	Electrical /Electronic Property Problem		
									A0705	Battery Problem		
							F19	Surgical Intervention	A07	Electrical /Electronic Property Problem		
									A0705	Battery Problem		
							F1905	Device Revision or Replacement	A07	Electrical /Electronic Property Problem		
									A0705	Battery Problem		
							F23	Unexpected Medical Intervention	A07	Electrical /Electronic Property Problem		
									A0705	Battery Problem		
						CT2985	Active	E011902	Presyncope	A07	Electrical /Electronic Property Problem		
										A0705	Battery Problem		
								E012207	Twitching	A07	Electrical /Electronic Property Problem		
										A0705	Battery Problem		
								E06	Heart	A07	Electrical /Electronic Property Problem		
										A0705	Battery Problem		
								F01	Change in Therapeutic Response	A07	Electrical /Electronic Property Problem		
										A0705	Battery Problem		
								F19	Surgical Intervention	A07	Electrical /Electronic Property Problem		
										A0705	Battery Problem		
								F1905	Device Revision or Replacement	A07	Electrical /Electronic Property Problem		
										A0705	Battery Problem		
								F23	Unexpected Medical Intervention	A07	Electrical /Electronic Property Problem		
										A0705	Battery Problem		
						CT3001	Implantable metal	E011902	Presyncope	A07	Electrical /Electronic Property Problem		
										A0705	Battery Problem		
								E012207	Twitching	A07	Electrical /Electronic Property Problem		
										A0705	Battery Problem		
								E06	Heart	A07	Electrical /Electronic Property Problem		
										A0705	Battery Problem		
								F01	Change in Therapeutic Response	A07	Electrical /Electronic Property Problem		
										A0705	Battery Problem		
								F19	Surgical Intervention	A07	Electrical /Electronic Property Problem		
										A0705	Battery Problem		
								F1905	Device Revision or Replacement	A07	Electrical /Electronic Property Problem		
										A0705	Battery Problem		
								F23	Unexpected Medical Intervention	A07	Electrical /Electronic Property Problem		
										A0705	Battery Problem		
						CT3002	Implantable synthetic polymer	E011902	Presyncope	A07	Electrical /Electronic Property Problem		
										A0705	Battery Problem		
								E012207	Twitching	A07	Electrical /Electronic Property Problem		
										A0705	Battery Problem		
								E06	Heart	A07	Electrical /Electronic Property Problem		
										A0705	Battery Problem		
								F01	Change in Therapeutic Response	A07	Electrical /Electronic Property Problem		
										A0705	Battery Problem		
								F19	Surgical Intervention	A07	Electrical /Electronic Property Problem		
										A0705	Battery Problem		
								F1905	Device Revision or Replacement	A07	Electrical /Electronic Property Problem		
										A0705	Battery Problem		
								F23	Unexpected Medical Intervention	A07	Electrical /Electronic Property Problem		
										A0705	Battery Problem		
						CT3006	Implantable inorganic material	E011902	Presyncope	A07	Electrical /Electronic Property Problem		
										A0705	Battery Problem		
								E012207	Twitching	A07	Electrical /Electronic Property Problem		
										A0705	Battery Problem		
								E06	Heart	A07	Electrical /Electronic Property Problem		
										A0705	Battery Problem		
								F01	Change in Therapeutic Response	A07	Electrical /Electronic Property Problem		
										A0705	Battery Problem		
								F19	Surgical Intervention	A07	Electrical /Electronic Property Problem		
										A0705	Battery Problem		
								F1905	Device Revision or Replacement	A07	Electrical /Electronic Property Problem		
										A0705	Battery Problem		
								F23	Unexpected Medical Intervention	A07	Electrical /Electronic Property Problem		
										A0705	Battery Problem		
						CT314	Electrically powered	E011902	Presyncope	A07	Electrical /Electronic Property Problem		
										A0705	Battery Problem		
								E012207	Twitching	A07	Electrical /Electronic Property Problem		
										A0705	Battery Problem		
								E06	Heart	A07	Electrical /Electronic Property Problem		
										A0705	Battery Problem		
								F01	Change in Therapeutic Response	A07	Electrical /Electronic Property Problem		
										A0705	Battery Problem		
								F19	Surgical Intervention	A07	Electrical /Electronic Property Problem		
										A0705	Battery Problem		
								F1905	Device Revision or Replacement	A07	Electrical /Electronic Property Problem		
										A0705	Battery Problem		
								F23	Unexpected Medical Intervention	A07	Electrical /Electronic Property Problem		
										A0705	Battery Problem		
						CT321	Long-term surgical invasive	E011902	Presyncope	A07	Electrical /Electronic Property Problem		
										A0705	Battery Problem		
								E012207	Twitching	A07	Electrical /Electronic Property Problem		
										A0705	Battery Problem		
								E06	Heart	A07	Electrical /Electronic Property Problem		
										A0705	Battery Problem		

							F01	Change in Therapeutic Response	A07	Electrical /Electronic Property Problem		
									A0705	Battery Problem		
							F19	Surgical Intervention	A07	Electrical /Electronic Property Problem		
									A0705	Battery Problem		
							F1905	Device Revision or Replacement	A07	Electrical /Electronic Property Problem		
									A0705	Battery Problem		
							F23	Unexpected Medical Intervention	A07	Electrical /Electronic Property Problem		
									A0705	Battery Problem		
					CT449	Generators and associated devices	E011902	Presyncope	A07	Electrical /Electronic Property Problem		
									A0705	Battery Problem		
							E012207	Twitching	A07	Electrical /Electronic Property Problem		
									A0705	Battery Problem		
							E06	Heart	A07	Electrical /Electronic Property Problem		
									A0705	Battery Problem		
							F01	Change in Therapeutic Response	A07	Electrical /Electronic Property Problem		
									A0705	Battery Problem		
							F19	Surgical Intervention	A07	Electrical /Electronic Property Problem		
									A0705	Battery Problem		
							F1905	Device Revision or Replacement	A07	Electrical /Electronic Property Problem		
									A0705	Battery Problem		
							F23	Unexpected Medical Intervention	A07	Electrical /Electronic Property Problem		
									A0705	Battery Problem		
					CT981	Single use	E011902	Presyncope	A07	Electrical /Electronic Property Problem		
									A0705	Battery Problem		
							E012207	Twitching	A07	Electrical /Electronic Property Problem		
									A0705	Battery Problem		
							E06	Heart	A07	Electrical /Electronic Property Problem		
									A0705	Battery Problem		
							F01	Change in Therapeutic Response	A07	Electrical /Electronic Property Problem		
									A0705	Battery Problem		
							F19	Surgical Intervention	A07	Electrical /Electronic Property Problem		
									A0705	Battery Problem		
							F1905	Device Revision or Replacement	A07	Electrical /Electronic Property Problem		
									A0705	Battery Problem		
							F23	Unexpected Medical Intervention	A07	Electrical /Electronic Property Problem		
									A0705	Battery Problem		
					CT983	Surgical invasive	E011902	Presyncope	A07	Electrical /Electronic Property Problem		
									A0705	Battery Problem		
							E012207	Twitching	A07	Electrical /Electronic Property Problem		
									A0705	Battery Problem		
							E06	Heart	A07	Electrical /Electronic Property Problem		
									A0705	Battery Problem		
							F01	Change in Therapeutic Response	A07	Electrical /Electronic Property Problem		
									A0705	Battery Problem		
							F19	Surgical Intervention	A07	Electrical /Electronic Property Problem		
									A0705	Battery Problem		
							F1905	Device Revision or Replacement	A07	Electrical /Electronic Property Problem		
									A0705	Battery Problem		
							F23	Unexpected Medical Intervention	A07	Electrical /Electronic Property Problem		
									A0705	Battery Problem		
14214	2023	All other reportable incidents	47263	Cardiac resynchronization therapy implantable pacemaker	CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070909	Oversensing		
							F26	No Health Consequences or Impact	A070909	Oversensing		
					CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A070909	Oversensing		
							F26	No Health Consequences or Impact	A070909	Oversensing		
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A070909	Oversensing		
							F26	No Health Consequences or Impact	A070909	Oversensing		
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A070909	Oversensing		
							F26	No Health Consequences or Impact	A070909	Oversensing		
14215	2023	All other reportable incidents	60789	Intracardiac pacemaker	CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT2818	Catheterization/Cannulation/Fluid path devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
14216	2023	All other reportable incidents	37265	Dual-chamber implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E2311	Discomfort	A070909	Oversensing		
									A071302	Inappropriate/Inadequate Shock/Stimulation		
							F12	Serious Injury/ Illness/ Impairment	A070909	Oversensing		
									A071302	Inappropriate/Inadequate Shock/Stimulation		
14217	2023	All other reportable incidents	47265	Dual-chamber implantable pacemaker, rate-responsive	CT1685	Cardiac pacemakers	E0601	Arrhythmia	A071102	Failure to Interrogate	05/10/2023	05/10/2023
									A071204	No Pacing	05/10/2023	05/10/2023
							F1901	Additional Surgery	A071102	Failure to Interrogate	05/10/2023	05/10/2023
									A071204	No Pacing	05/10/2023	05/10/2023
					CT257	Implantable cardiac pacemakers	E0601	Arrhythmia	A071102	Failure to Interrogate	05/10/2023	05/10/2023
									A071204	No Pacing	05/10/2023	05/10/2023
							F1901	Additional Surgery	A071102	Failure to Interrogate	05/10/2023	05/10/2023
									A071204	No Pacing	05/10/2023	05/10/2023
					CT2866	Implantable cardiac electrical stimulators	E0601	Arrhythmia	A071102	Failure to Interrogate	05/10/2023	05/10/2023
									A071204	No Pacing	05/10/2023	05/10/2023
							F1901	Additional Surgery	A071102	Failure to Interrogate	05/10/2023	05/10/2023
									A071204	No Pacing	05/10/2023	05/10/2023
14218	2023	All other reportable incidents	47265	Dual-chamber implantable pacemaker, rate-responsive	CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A071102	Failure to Interrogate	20/07/2023	20/07/2023
							F1901	Additional Surgery	A071102	Failure to Interrogate	20/07/2023	20/07/2023
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A071102	Failure to Interrogate	20/07/2023	20/07/2023
							F1901	Additional Surgery	A071102	Failure to Interrogate	20/07/2023	20/07/2023
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A071102	Failure to Interrogate	20/07/2023	20/07/2023
							F1901	Additional Surgery	A071102	Failure to Interrogate	20/07/2023	20/07/2023

14219	2023	All other reportable incidents	47267	Single-chamber implantable pacemaker, rate-responsive	CT141	Electrodes/Leads and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	31/05/2023	31/05/2023
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	31/05/2023	31/05/2023
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	31/05/2023	31/05/2023
					CT145	Cardiovascular devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	31/05/2023	31/05/2023
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	31/05/2023	31/05/2023
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	31/05/2023	31/05/2023
					CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	31/05/2023	31/05/2023
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	31/05/2023	31/05/2023
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	31/05/2023	31/05/2023
					CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	31/05/2023	31/05/2023
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	31/05/2023	31/05/2023
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	31/05/2023	31/05/2023
					CT233	Surgical	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	31/05/2023	31/05/2023
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	31/05/2023	31/05/2023
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	31/05/2023	31/05/2023
					CT2406	Implantable	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	31/05/2023	31/05/2023
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	31/05/2023	31/05/2023
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	31/05/2023	31/05/2023
					CT2407	Active implantable	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	31/05/2023	31/05/2023
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	31/05/2023	31/05/2023
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	31/05/2023	31/05/2023
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	31/05/2023	31/05/2023
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	31/05/2023	31/05/2023
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	31/05/2023	31/05/2023
					CT2727	Active implant devices/ancillaries	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	31/05/2023	31/05/2023
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	31/05/2023	31/05/2023
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	31/05/2023	31/05/2023
					CT2750	Implantable cardiac electrical stimulators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	31/05/2023	31/05/2023
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	31/05/2023	31/05/2023
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	31/05/2023	31/05/2023
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	31/05/2023	31/05/2023
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	31/05/2023	31/05/2023
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	31/05/2023	31/05/2023
					CT2985	Active	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	31/05/2023	31/05/2023
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	31/05/2023	31/05/2023
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	31/05/2023	31/05/2023
					CT3001	Implantable metal	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	31/05/2023	31/05/2023
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	31/05/2023	31/05/2023
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	31/05/2023	31/05/2023
					CT3002	Implantable synthetic polymer	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	31/05/2023	31/05/2023
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	31/05/2023	31/05/2023
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	31/05/2023	31/05/2023
					CT3006	Implantable inorganic material	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	31/05/2023	31/05/2023
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	31/05/2023	31/05/2023
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	31/05/2023	31/05/2023
					CT314	Electrically powered	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	31/05/2023	31/05/2023
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	31/05/2023	31/05/2023
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	31/05/2023	31/05/2023
					CT321	Long-term surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	31/05/2023	31/05/2023
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	31/05/2023	31/05/2023
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	31/05/2023	31/05/2023
					CT449	Generators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	31/05/2023	31/05/2023
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	31/05/2023	31/05/2023
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	31/05/2023	31/05/2023
					CT981	Single use	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	31/05/2023	31/05/2023
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	31/05/2023	31/05/2023
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	31/05/2023	31/05/2023
					CT983	Surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	31/05/2023	31/05/2023
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	31/05/2023	31/05/2023
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	31/05/2023	31/05/2023
14220	2023	All other reportable incidents	47270	Cardiac resynchronization therapy implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A070909	Oversensing		
									A090801	Signal Artifact/Noise		
							F08	Hospitalization or Prolonged Hospitalization	A070909	Oversensing		
									A090801	Signal Artifact/Noise		
							F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A070909	Oversensing		
									A090801	Signal Artifact/Noise		
							F1901	Additional Surgery	A070909	Oversensing		
									A090801	Signal Artifact/Noise		
							F23	Unexpected Medical Intervention	A070909	Oversensing		
									A090801	Signal Artifact/Noise		
							F2301	Additional Device Required	A070909	Oversensing		
									A090801	Signal Artifact/Noise		
14221	2023	All other reportable incidents	47267	Single-chamber implantable pacemaker, rate-responsive	CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
14222	2023	All other reportable incidents	17882	Physiologic-monitoring defibrillation system	CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070803	Failure to Power Up		
							F27	Problem Identified During Non-Clinical Procedure	A070803	Failure to Power Up		
14223	2023	All other reportable incidents	47267	Single-chamber implantable pacemaker, rate-responsive	CT141	Electrodes/Leads and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070909	Oversensing		
									A072201	High Impedance		
									A072202	Low Impedance		
									A090801	Signal Artifact/Noise		
									A1301	Failure to Read Input Signal		
									A1601	Device Alarm System		
							F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A070909	Oversensing		
									A072201	High Impedance		

								A072202	Low Impedance		
								A090801	Signal Artifact/Noise		
								A1301	Failure to Read Input Signal		
								A1601	Device Alarm System		
					CT145	Cardiovascular devices	E2403	No Clinical Signs, Symptoms or Conditions	A070909	Oversensing		
									A072201	High Impedance		
									A072202	Low Impedance		
									A090801	Signal Artifact/Noise		
									A1301	Failure to Read Input Signal		
									A1601	Device Alarm System		
							F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A070909	Oversensing		
									A072201	High Impedance		
									A072202	Low Impedance		
									A090801	Signal Artifact/Noise		
									A1301	Failure to Read Input Signal		
									A1601	Device Alarm System		
					CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070909	Oversensing		
									A072201	High Impedance		
									A072202	Low Impedance		
									A090801	Signal Artifact/Noise		
									A1301	Failure to Read Input Signal		
									A1601	Device Alarm System		
							F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A070909	Oversensing		
									A072201	High Impedance		
									A072202	Low Impedance		
									A090801	Signal Artifact/Noise		
									A1301	Failure to Read Input Signal		
									A1601	Device Alarm System		
					CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A070909	Oversensing		
									A072201	High Impedance		
									A072202	Low Impedance		
									A090801	Signal Artifact/Noise		
									A1301	Failure to Read Input Signal		
									A1601	Device Alarm System		
							F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A070909	Oversensing		
									A072201	High Impedance		
									A072202	Low Impedance		
									A090801	Signal Artifact/Noise		
									A1301	Failure to Read Input Signal		
									A1601	Device Alarm System		
					CT233	Surgical	E2403	No Clinical Signs, Symptoms or Conditions	A070909	Oversensing		
									A072201	High Impedance		
									A072202	Low Impedance		
									A090801	Signal Artifact/Noise		
									A1301	Failure to Read Input Signal		
									A1601	Device Alarm System		
							F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A070909	Oversensing		
									A072201	High Impedance		
									A072202	Low Impedance		
									A090801	Signal Artifact/Noise		
									A1301	Failure to Read Input Signal		
									A1601	Device Alarm System		
					CT2406	Implantable	E2403	No Clinical Signs, Symptoms or Conditions	A070909	Oversensing		
									A072201	High Impedance		
									A072202	Low Impedance		
									A090801	Signal Artifact/Noise		
									A1301	Failure to Read Input Signal		
									A1601	Device Alarm System		
							F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A070909	Oversensing		
									A072201	High Impedance		
									A072202	Low Impedance		
									A090801	Signal Artifact/Noise		
									A1301	Failure to Read Input Signal		
									A1601	Device Alarm System		
					CT2407	Active implantable	E2403	No Clinical Signs, Symptoms or Conditions	A070909	Oversensing		
									A072201	High Impedance		
									A072202	Low Impedance		
									A090801	Signal Artifact/Noise		
									A1301	Failure to Read Input Signal		
									A1601	Device Alarm System		
							F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A070909	Oversensing		
									A072201	High Impedance		
									A072202	Low Impedance		
									A090801	Signal Artifact/Noise		
									A1301	Failure to Read Input Signal		
									A1601	Device Alarm System		
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A070909	Oversensing		
									A072201	High Impedance		
									A072202	Low Impedance		
									A090801	Signal Artifact/Noise		
									A1301	Failure to Read Input Signal		
									A1601	Device Alarm System		
							F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A070909	Oversensing		
									A072201	High Impedance		
									A072202	Low Impedance		
									A090801	Signal Artifact/Noise		
									A1301	Failure to Read Input Signal		
									A1601	Device Alarm System		

					CT2727	Active implant devices/ancillaries	E2403	No Clinical Signs, Symptoms or Conditions	A070909	Oversensing		
									A072201	High Impedance		
									A072202	Low Impedance		
									A090801	Signal Artifact/Noise		
									A1301	Failure to Read Input Signal		
									A1601	Device Alarm System		
							F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A070909	Oversensing		
									A072201	High Impedance		
									A072202	Low Impedance		
									A090801	Signal Artifact/Noise		
									A1301	Failure to Read Input Signal		
									A1601	Device Alarm System		
					CT2750	Implantable cardiac electrical stimulators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070909	Oversensing		
									A072201	High Impedance		
									A072202	Low Impedance		
									A090801	Signal Artifact/Noise		
									A1301	Failure to Read Input Signal		
									A1601	Device Alarm System		
							F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A070909	Oversensing		
									A072201	High Impedance		
									A072202	Low Impedance		
									A090801	Signal Artifact/Noise		
									A1301	Failure to Read Input Signal		
									A1601	Device Alarm System		
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A070909	Oversensing		
									A072201	High Impedance		
									A072202	Low Impedance		
									A090801	Signal Artifact/Noise		
									A1301	Failure to Read Input Signal		
									A1601	Device Alarm System		
							F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A070909	Oversensing		
									A072201	High Impedance		
									A072202	Low Impedance		
									A090801	Signal Artifact/Noise		
									A1301	Failure to Read Input Signal		
									A1601	Device Alarm System		
					CT2985	Active	E2403	No Clinical Signs, Symptoms or Conditions	A070909	Oversensing		
									A072201	High Impedance		
									A072202	Low Impedance		
									A090801	Signal Artifact/Noise		
									A1301	Failure to Read Input Signal		
									A1601	Device Alarm System		
							F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A070909	Oversensing		
									A072201	High Impedance		
									A072202	Low Impedance		
									A090801	Signal Artifact/Noise		
									A1301	Failure to Read Input Signal		
									A1601	Device Alarm System		
					CT3001	Implantable metal	E2403	No Clinical Signs, Symptoms or Conditions	A070909	Oversensing		
									A072201	High Impedance		
									A072202	Low Impedance		
									A090801	Signal Artifact/Noise		
									A1301	Failure to Read Input Signal		
									A1601	Device Alarm System		
							F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A070909	Oversensing		
									A072201	High Impedance		
									A072202	Low Impedance		
									A090801	Signal Artifact/Noise		
									A1301	Failure to Read Input Signal		
									A1601	Device Alarm System		
					CT3002	Implantable synthetic polymer	E2403	No Clinical Signs, Symptoms or Conditions	A070909	Oversensing		
									A072201	High Impedance		
									A072202	Low Impedance		
									A090801	Signal Artifact/Noise		
									A1301	Failure to Read Input Signal		
									A1601	Device Alarm System		
							F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A070909	Oversensing		
									A072201	High Impedance		
									A072202	Low Impedance		
									A090801	Signal Artifact/Noise		
									A1301	Failure to Read Input Signal		
									A1601	Device Alarm System		
					CT3006	Implantable inorganic material	E2403	No Clinical Signs, Symptoms or Conditions	A070909	Oversensing		
									A072201	High Impedance		
									A072202	Low Impedance		
									A090801	Signal Artifact/Noise		
									A1301	Failure to Read Input Signal		
									A1601	Device Alarm System		
							F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A070909	Oversensing		
									A072201	High Impedance		
									A072202	Low Impedance		
									A090801	Signal Artifact/Noise		
									A1301	Failure to Read Input Signal		
									A1601	Device Alarm System		
					CT314	Electrically powered	E2403	No Clinical Signs, Symptoms or Conditions	A070909	Oversensing		
									A072201	High Impedance		
									A072202	Low Impedance		
									A090801	Signal Artifact/Noise		

									A1301	Failure to Read Input Signal		
									A1601	Device Alarm System		
							F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A070909	Oversensing		
									A072201	High Impedance		
									A072202	Low Impedance		
									A090801	Signal Artifact/Noise		
									A1301	Failure to Read Input Signal		
									A1601	Device Alarm System		
					CT321	Long-term surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A070909	Oversensing		
									A072201	High Impedance		
									A072202	Low Impedance		
									A090801	Signal Artifact/Noise		
									A1301	Failure to Read Input Signal		
									A1601	Device Alarm System		
							F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A070909	Oversensing		
									A072201	High Impedance		
									A072202	Low Impedance		
									A090801	Signal Artifact/Noise		
									A1301	Failure to Read Input Signal		
									A1601	Device Alarm System		
					CT449	Generators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070909	Oversensing		
									A072201	High Impedance		
									A072202	Low Impedance		
									A090801	Signal Artifact/Noise		
									A1301	Failure to Read Input Signal		
									A1601	Device Alarm System		
							F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A070909	Oversensing		
									A072201	High Impedance		
									A072202	Low Impedance		
									A090801	Signal Artifact/Noise		
									A1301	Failure to Read Input Signal		
									A1601	Device Alarm System		
					CT981	Single use	E2403	No Clinical Signs, Symptoms or Conditions	A070909	Oversensing		
									A072201	High Impedance		
									A072202	Low Impedance		
									A090801	Signal Artifact/Noise		
									A1301	Failure to Read Input Signal		
									A1601	Device Alarm System		
							F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A070909	Oversensing		
									A072201	High Impedance		
									A072202	Low Impedance		
									A090801	Signal Artifact/Noise		
									A1301	Failure to Read Input Signal		
									A1601	Device Alarm System		
					CT983	Surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A070909	Oversensing		
									A072201	High Impedance		
									A072202	Low Impedance		
									A090801	Signal Artifact/Noise		
									A1301	Failure to Read Input Signal		
									A1601	Device Alarm System		
							F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A070909	Oversensing		
									A072201	High Impedance		
									A072202	Low Impedance		
									A090801	Signal Artifact/Noise		
									A1301	Failure to Read Input Signal		
									A1601	Device Alarm System		
14224	2023	All other reportable incidents	35852	Single-chamber implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E2104	Electric Shock	A070909	Oversensing		
									A071302	Inappropriate/Inadequate Shock/Stimulation		
							F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A070909	Oversensing		
									A071302	Inappropriate/Inadequate Shock/Stimulation		
14225	2023	All other reportable incidents	36241	Pacemaker/defibrillator lead	CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070909	Oversensing		
									A072201	High Impedance		
									A072202	Low Impedance		
									A090801	Signal Artifact/Noise		
									A1301	Failure to Read Input Signal		
							F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A070909	Oversensing		
									A072201	High Impedance		
									A072202	Low Impedance		
									A090801	Signal Artifact/Noise		
									A1301	Failure to Read Input Signal		
14226	2023	All other reportable incidents	47265	Dual-chamber implantable pacemaker, rate-responsive	CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
14227	2023		47265	Dual-chamber implantable pacemaker, rate-responsive	CT141	Electrodes/Leads and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	24/09/2023	24/09/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	24/09/2023	24/09/2023
					CT145	Cardiovascular devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	24/09/2023	24/09/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	24/09/2023	24/09/2023
					CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	24/09/2023	24/09/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	24/09/2023	24/09/2023
					CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	24/09/2023	24/09/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	24/09/2023	24/09/2023
					CT233	Surgical	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	24/09/2023	24/09/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	24/09/2023	24/09/2023
					CT2406	Implantable	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	24/09/2023	24/09/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	24/09/2023	24/09/2023

					CT2407	Active implantable	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	24/09/2023	24/09/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	24/09/2023	24/09/2023
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	24/09/2023	24/09/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	24/09/2023	24/09/2023
					CT2727	Active implant devices/ancillaries	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	24/09/2023	24/09/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	24/09/2023	24/09/2023
					CT2750	Implantable cardiac electrical stimulators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	24/09/2023	24/09/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	24/09/2023	24/09/2023
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	24/09/2023	24/09/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	24/09/2023	24/09/2023
					CT2985	Active	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	24/09/2023	24/09/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	24/09/2023	24/09/2023
					CT3001	Implantable metal	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	24/09/2023	24/09/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	24/09/2023	24/09/2023
					CT3002	Implantable synthetic polymer	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	24/09/2023	24/09/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	24/09/2023	24/09/2023
					CT3006	Implantable inorganic material	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	24/09/2023	24/09/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	24/09/2023	24/09/2023
					CT314	Electrically powered	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	24/09/2023	24/09/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	24/09/2023	24/09/2023
					CT321	Long-term surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	24/09/2023	24/09/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	24/09/2023	24/09/2023
					CT449	Generators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	24/09/2023	24/09/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	24/09/2023	24/09/2023
					CT981	Single use	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	24/09/2023	24/09/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	24/09/2023	24/09/2023
					CT983	Surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	24/09/2023	24/09/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	24/09/2023	24/09/2023
14228	2023		47265	Dual-chamber implantable pacemaker, rate-responsive	CT141	Electrodes/Leads and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	18/08/2023	18/08/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	18/08/2023	18/08/2023
					CT145	Cardiovascular devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	18/08/2023	18/08/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	18/08/2023	18/08/2023
					CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	18/08/2023	18/08/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	18/08/2023	18/08/2023
					CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	18/08/2023	18/08/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	18/08/2023	18/08/2023
					CT233	Surgical	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	18/08/2023	18/08/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	18/08/2023	18/08/2023
					CT2406	Implantable	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	18/08/2023	18/08/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	18/08/2023	18/08/2023
					CT2407	Active implantable	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	18/08/2023	18/08/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	18/08/2023	18/08/2023
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	18/08/2023	18/08/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	18/08/2023	18/08/2023
					CT2727	Active implant devices/ancillaries	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	18/08/2023	18/08/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	18/08/2023	18/08/2023
					CT2750	Implantable cardiac electrical stimulators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	18/08/2023	18/08/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	18/08/2023	18/08/2023
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	18/08/2023	18/08/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	18/08/2023	18/08/2023
					CT2985	Active	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	18/08/2023	18/08/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	18/08/2023	18/08/2023
					CT3001	Implantable metal	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	18/08/2023	18/08/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	18/08/2023	18/08/2023
					CT3002	Implantable synthetic polymer	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	18/08/2023	18/08/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	18/08/2023	18/08/2023
					CT3006	Implantable inorganic material	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	18/08/2023	18/08/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	18/08/2023	18/08/2023
					CT314	Electrically powered	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	18/08/2023	18/08/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	18/08/2023	18/08/2023
					CT321	Long-term surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	18/08/2023	18/08/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	18/08/2023	18/08/2023
					CT449	Generators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	18/08/2023	18/08/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	18/08/2023	18/08/2023
					CT981	Single use	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	18/08/2023	18/08/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	18/08/2023	18/08/2023
					CT983	Surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	18/08/2023	18/08/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	18/08/2023	18/08/2023
14229	2023		47265	Dual-chamber implantable pacemaker, rate-responsive	CT141	Electrodes/Leads and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	14/10/2023	14/10/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	14/10/2023	14/10/2023
					CT145	Cardiovascular devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	14/10/2023	14/10/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	14/10/2023	14/10/2023
					CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	14/10/2023	14/10/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	14/10/2023	14/10/2023
					CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	14/10/2023	14/10/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	14/10/2023	14/10/2023
					CT233	Surgical	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	14/10/2023	14/10/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	14/10/2023	14/10/2023
					CT2406	Implantable	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	14/10/2023	14/10/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	14/10/2023	14/10/2023
					CT2407	Active implantable	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	14/10/2023	14/10/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	14/10/2023	14/10/2023
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	14/10/2023	14/10/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	14/10/2023	14/10/2023
					CT2727	Active implant devices/ancillaries	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	14/10/2023	14/10/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	14/10/2023	14/10/2023
					CT2750	Implantable cardiac electrical stimulators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	14/10/2023	14/10/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	14/10/2023	14/10/2023

					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	14/10/2023	14/10/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	14/10/2023	14/10/2023
					CT2985	Active	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	14/10/2023	14/10/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	14/10/2023	14/10/2023
					CT3001	Implantable metal	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	14/10/2023	14/10/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	14/10/2023	14/10/2023
					CT3002	Implantable synthetic polymer	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	14/10/2023	14/10/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	14/10/2023	14/10/2023
					CT3006	Implantable inorganic material	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	14/10/2023	14/10/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	14/10/2023	14/10/2023
					CT314	Electrically powered	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	14/10/2023	14/10/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	14/10/2023	14/10/2023
					CT321	Long-term surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	14/10/2023	14/10/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	14/10/2023	14/10/2023
					CT449	Generators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	14/10/2023	14/10/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	14/10/2023	14/10/2023
					CT981	Single use	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	14/10/2023	14/10/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	14/10/2023	14/10/2023
					CT983	Surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	14/10/2023	14/10/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	14/10/2023	14/10/2023
14230	2023		47265	Dual-chamber implantable pacemaker, rate-responsive	CT141	Electrodes/Leads and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	12/10/2023	12/10/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	12/10/2023	12/10/2023
					CT145	Cardiovascular devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	12/10/2023	12/10/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	12/10/2023	12/10/2023
					CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	12/10/2023	12/10/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	12/10/2023	12/10/2023
					CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	12/10/2023	12/10/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	12/10/2023	12/10/2023
					CT233	Surgical	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	12/10/2023	12/10/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	12/10/2023	12/10/2023
					CT2406	Implantable	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	12/10/2023	12/10/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	12/10/2023	12/10/2023
					CT2407	Active implantable	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	12/10/2023	12/10/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	12/10/2023	12/10/2023
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	12/10/2023	12/10/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	12/10/2023	12/10/2023
					CT2727	Active implant devices/ancillaries	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	12/10/2023	12/10/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	12/10/2023	12/10/2023
					CT2750	Implantable cardiac electrical stimulators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	12/10/2023	12/10/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	12/10/2023	12/10/2023
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	12/10/2023	12/10/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	12/10/2023	12/10/2023
					CT2985	Active	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	12/10/2023	12/10/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	12/10/2023	12/10/2023
					CT3001	Implantable metal	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	12/10/2023	12/10/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	12/10/2023	12/10/2023
					CT3002	Implantable synthetic polymer	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	12/10/2023	12/10/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	12/10/2023	12/10/2023
					CT3006	Implantable inorganic material	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	12/10/2023	12/10/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	12/10/2023	12/10/2023
					CT314	Electrically powered	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	12/10/2023	12/10/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	12/10/2023	12/10/2023
					CT321	Long-term surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	12/10/2023	12/10/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	12/10/2023	12/10/2023
					CT449	Generators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	12/10/2023	12/10/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	12/10/2023	12/10/2023
					CT981	Single use	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	12/10/2023	12/10/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	12/10/2023	12/10/2023
					CT983	Surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	12/10/2023	12/10/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	12/10/2023	12/10/2023
14231	2023	All other reportable incidents	47265	Dual-chamber implantable pacemaker, rate-responsive	CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A05	Mechanical Problem	16/10/2023	16/10/2023
							F26	No Health Consequences or Impact	A05	Mechanical Problem	16/10/2023	16/10/2023
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A05	Mechanical Problem	16/10/2023	16/10/2023
							F26	No Health Consequences or Impact	A05	Mechanical Problem	16/10/2023	16/10/2023
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A05	Mechanical Problem	16/10/2023	16/10/2023
							F26	No Health Consequences or Impact	A05	Mechanical Problem	16/10/2023	16/10/2023
14232	2023	All other reportable incidents	63188	Left-ventricular wireless pacing system sensor/transmitter	CT141	Electrodes/Leads and associated devices			A0708	Power Problem		
					CT145	Cardiovascular devices			A0708	Power Problem		
					CT1577	Wireless monitoring system transmitters			A0708	Power Problem		
					CT1684	Cardiac pacemakers and associated devices			A0708	Power Problem		
					CT2317	Transmitters/Receivers/Transceivers			A0708	Power Problem		
					CT233	Surgical			A0708	Power Problem		
					CT2406	Implantable			A0708	Power Problem		
					CT2407	Active implantable			A0708	Power Problem		
					CT255	Transducers and associated devices			A0708	Power Problem		
					CT2726	Physiological monitoring/measuring devices			A0708	Power Problem		
					CT2727	Active implant devices/ancillaries			A0708	Power Problem		
					CT2750	Implantable cardiac electrical stimulators and associated devices			A0708	Power Problem		
					CT2885	Physiological monitoring/measuring system transmitters/receivers/transceivers			A0708	Power Problem		
					CT2985	Active			A0708	Power Problem		
					CT3001	Implantable metal			A0708	Power Problem		
					CT3002	Implantable synthetic polymer			A0708	Power Problem		
					CT3006	Implantable inorganic material			A0708	Power Problem		
					CT314	Electrically powered			A0708	Power Problem		
					CT321	Long-term surgical invasive			A0708	Power Problem		
					CT384	Sensors and associated devices			A0708	Power Problem		

					CT449	Generators and associated devices			A0708	Power Problem		
					CT981	Single use			A0708	Power Problem		
					CT983	Surgical invasive			A0708	Power Problem		
					CT994	Sensors			A0708	Power Problem		
14233	2023		47265	Dual-chamber implantable pacemaker, rate-responsive	CT141	Electrodes/Leads and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	11/10/2023	11/10/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	11/10/2023	11/10/2023
					CT145	Cardiovascular devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	11/10/2023	11/10/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	11/10/2023	11/10/2023
					CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	11/10/2023	11/10/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	11/10/2023	11/10/2023
					CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	11/10/2023	11/10/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	11/10/2023	11/10/2023
					CT233	Surgical	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	11/10/2023	11/10/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	11/10/2023	11/10/2023
					CT2406	Implantable	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	11/10/2023	11/10/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	11/10/2023	11/10/2023
					CT2407	Active implantable	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	11/10/2023	11/10/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	11/10/2023	11/10/2023
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	11/10/2023	11/10/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	11/10/2023	11/10/2023
					CT2727	Active implant devices/ancillaries	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	11/10/2023	11/10/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	11/10/2023	11/10/2023
					CT2750	Implantable cardiac electrical stimulators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	11/10/2023	11/10/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	11/10/2023	11/10/2023
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	11/10/2023	11/10/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	11/10/2023	11/10/2023
					CT2985	Active	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	11/10/2023	11/10/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	11/10/2023	11/10/2023
					CT3001	Implantable metal	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	11/10/2023	11/10/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	11/10/2023	11/10/2023
					CT3002	Implantable synthetic polymer	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	11/10/2023	11/10/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	11/10/2023	11/10/2023
					CT3006	Implantable inorganic material	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	11/10/2023	11/10/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	11/10/2023	11/10/2023
					CT314	Electrically powered	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	11/10/2023	11/10/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	11/10/2023	11/10/2023
					CT321	Long-term surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	11/10/2023	11/10/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	11/10/2023	11/10/2023
					CT449	Generators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	11/10/2023	11/10/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	11/10/2023	11/10/2023
					CT981	Single use	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	11/10/2023	11/10/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	11/10/2023	11/10/2023
					CT983	Surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	11/10/2023	11/10/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	11/10/2023	11/10/2023
14234	2023	All other reportable incidents	36241	Pacemaker/defibrillator lead	CT1684	Cardiac pacemakers and associated devices	E2104	Electric Shock	A0709	Device Sensing Problem	10/01/2024	10/01/2024
									A070909	Oversensing	10/01/2024	10/01/2024
									A071302	Inappropriate/Inadequate Shock/Stimulation	10/01/2024	10/01/2024
							F08	Hospitalization or Prolonged Hospitalization	A0709	Device Sensing Problem	10/01/2024	10/01/2024
									A070909	Oversensing	10/01/2024	10/01/2024
									A071302	Inappropriate/Inadequate Shock/Stimulation	10/01/2024	10/01/2024
							F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A0709	Device Sensing Problem	10/01/2024	10/01/2024
									A070909	Oversensing	10/01/2024	10/01/2024
									A071302	Inappropriate/Inadequate Shock/Stimulation	10/01/2024	10/01/2024
							F1903	Device Explantation	A0709	Device Sensing Problem	10/01/2024	10/01/2024
									A070909	Oversensing	10/01/2024	10/01/2024
									A071302	Inappropriate/Inadequate Shock/Stimulation	10/01/2024	10/01/2024
14235	2023	All other reportable incidents	36241	Pacemaker/defibrillator lead	CT1684	Cardiac pacemakers and associated devices	E2104	Electric Shock	A072201	High Impedance		
									A1301	Failure to Read Input Signal		
							F17	Sedation	A072201	High Impedance		
									A1301	Failure to Read Input Signal		
							F22	Unexpected Diagnostic Intervention	A072201	High Impedance		
									A1301	Failure to Read Input Signal		
14236	2023	All other reportable incidents	47265	Dual-chamber implantable pacemaker, rate-responsive	CT141	Electrodes/Leads and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	10/10/2023	10/10/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	10/10/2023	10/10/2023
					CT145	Cardiovascular devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	10/10/2023	10/10/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	10/10/2023	10/10/2023
					CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	10/10/2023	10/10/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	10/10/2023	10/10/2023
					CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	10/10/2023	10/10/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	10/10/2023	10/10/2023
					CT233	Surgical	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	10/10/2023	10/10/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	10/10/2023	10/10/2023
					CT2406	Implantable	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	10/10/2023	10/10/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	10/10/2023	10/10/2023
					CT2407	Active implantable	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	10/10/2023	10/10/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	10/10/2023	10/10/2023
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	10/10/2023	10/10/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	10/10/2023	10/10/2023
					CT2727	Active implant devices/ancillaries	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	10/10/2023	10/10/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	10/10/2023	10/10/2023
					CT2750	Implantable cardiac electrical stimulators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	10/10/2023	10/10/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	10/10/2023	10/10/2023
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	10/10/2023	10/10/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	10/10/2023	10/10/2023
					CT2985	Active	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	10/10/2023	10/10/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	10/10/2023	10/10/2023
					CT3001	Implantable metal	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	10/10/2023	10/10/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	10/10/2023	10/10/2023

					CT3002	Implantable synthetic polymer	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	10/10/2023	10/10/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	10/10/2023	10/10/2023
					CT3006	Implantable inorganic material	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	10/10/2023	10/10/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	10/10/2023	10/10/2023
					CT314	Electrically powered	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	10/10/2023	10/10/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	10/10/2023	10/10/2023
					CT321	Long-term surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	10/10/2023	10/10/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	10/10/2023	10/10/2023
					CT449	Generators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	10/10/2023	10/10/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	10/10/2023	10/10/2023
					CT981	Single use	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	10/10/2023	10/10/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	10/10/2023	10/10/2023
					CT983	Surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	10/10/2023	10/10/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	10/10/2023	10/10/2023
14237	2023	All other reportable incidents	35852	Single-chamber implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
14238	2023		47265	Dual-chamber implantable pacemaker, rate-responsive	CT141	Electrodes/Leads and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	14/10/2023	14/10/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	14/10/2023	14/10/2023
					CT145	Cardiovascular devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	14/10/2023	14/10/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	14/10/2023	14/10/2023
					CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	14/10/2023	14/10/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	14/10/2023	14/10/2023
					CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	14/10/2023	14/10/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	14/10/2023	14/10/2023
					CT233	Surgical	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	14/10/2023	14/10/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	14/10/2023	14/10/2023
					CT2406	Implantable	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	14/10/2023	14/10/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	14/10/2023	14/10/2023
					CT2407	Active implantable	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	14/10/2023	14/10/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	14/10/2023	14/10/2023
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	14/10/2023	14/10/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	14/10/2023	14/10/2023
					CT2727	Active implant devices/ancillaries	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	14/10/2023	14/10/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	14/10/2023	14/10/2023
					CT2750	Implantable cardiac electrical stimulators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	14/10/2023	14/10/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	14/10/2023	14/10/2023
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	14/10/2023	14/10/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	14/10/2023	14/10/2023
					CT2985	Active	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	14/10/2023	14/10/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	14/10/2023	14/10/2023
					CT3001	Implantable metal	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	14/10/2023	14/10/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	14/10/2023	14/10/2023
					CT3002	Implantable synthetic polymer	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	14/10/2023	14/10/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	14/10/2023	14/10/2023
					CT3006	Implantable inorganic material	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	14/10/2023	14/10/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	14/10/2023	14/10/2023
					CT314	Electrically powered	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	14/10/2023	14/10/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	14/10/2023	14/10/2023
					CT321	Long-term surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	14/10/2023	14/10/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	14/10/2023	14/10/2023
					CT449	Generators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	14/10/2023	14/10/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	14/10/2023	14/10/2023
					CT981	Single use	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	14/10/2023	14/10/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	14/10/2023	14/10/2023
					CT983	Surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	14/10/2023	14/10/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	14/10/2023	14/10/2023
14239	2023	All other reportable incidents	35223	Endocardial/interventricular septal pacing lead	CT1684	Cardiac pacemakers and associated devices	E0506	Hemorrhage/Blood Loss/Bleeding	A24	No Apparent Device Problem	09/10/2023	09/10/2023
							E0511	Perforation of Vessels	A24	No Apparent Device Problem	09/10/2023	09/10/2023
							E0605	Cardiac Tamponade	A24	No Apparent Device Problem	09/10/2023	09/10/2023
							F1908	Prolonged Surgery	A24	No Apparent Device Problem	09/10/2023	09/10/2023
14240	2023	All other reportable incidents	47265	Dual-chamber implantable pacemaker, rate-responsive	CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	26/02/2023	26/02/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	26/02/2023	26/02/2023
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	26/02/2023	26/02/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	26/02/2023	26/02/2023
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	26/02/2023	26/02/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	26/02/2023	26/02/2023
14241	2023		47265	Dual-chamber implantable pacemaker, rate-responsive	CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	27/09/2023	27/09/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	27/09/2023	27/09/2023
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	27/09/2023	27/09/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	27/09/2023	27/09/2023
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	27/09/2023	27/09/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	27/09/2023	27/09/2023
14242	2023	All other reportable incidents	47267	Single-chamber implantable pacemaker, rate-responsive	CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A0908	Incorrect, Inadequate or Imprecise Result or Readings		
							F26	No Health Consequences or Impact	A0908	Incorrect, Inadequate or Imprecise Result or Readings		
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A0908	Incorrect, Inadequate or Imprecise Result or Readings		
							F26	No Health Consequences or Impact	A0908	Incorrect, Inadequate or Imprecise Result or Readings		
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A0908	Incorrect, Inadequate or Imprecise Result or Readings		
							F26	No Health Consequences or Impact	A0908	Incorrect, Inadequate or Imprecise Result or Readings		
14243	2023	All other reportable incidents	47265	Dual-chamber implantable pacemaker, rate-responsive	CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		

							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
14244	2023	All other reportable incidents	47265	Dual-chamber implantable pacemaker, rate-responsive	CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery		
							F12	Serious Injury/ Illness/ Impairment	A070504	Premature Discharge of Battery		
							F1903	Device Explantation	A070504	Premature Discharge of Battery		
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery		
							F12	Serious Injury/ Illness/ Impairment	A070504	Premature Discharge of Battery		
							F1903	Device Explantation	A070504	Premature Discharge of Battery		
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery		
							F12	Serious Injury/ Illness/ Impairment	A070504	Premature Discharge of Battery		
							F1903	Device Explantation	A070504	Premature Discharge of Battery		
14245	2023	All other reportable incidents	47267	Single-chamber implantable pacemaker, rate-responsive	CT141	Electrodes/Leads and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A0908	Incorrect, Inadequate or Imprecise Result or Readings	16/10/2023	16/10/2023
							F08	Hospitalization or Prolonged Hospitalization	A0908	Incorrect, Inadequate or Imprecise Result or Readings	16/10/2023	16/10/2023
							F1905	Device Revision or Replacement	A0908	Incorrect, Inadequate or Imprecise Result or Readings	16/10/2023	16/10/2023
					CT145	Cardiovascular devices	E2403	No Clinical Signs, Symptoms or Conditions	A0908	Incorrect, Inadequate or Imprecise Result or Readings	16/10/2023	16/10/2023
							F08	Hospitalization or Prolonged Hospitalization	A0908	Incorrect, Inadequate or Imprecise Result or Readings	16/10/2023	16/10/2023
							F1905	Device Revision or Replacement	A0908	Incorrect, Inadequate or Imprecise Result or Readings	16/10/2023	16/10/2023
					CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A0908	Incorrect, Inadequate or Imprecise Result or Readings	16/10/2023	16/10/2023
							F08	Hospitalization or Prolonged Hospitalization	A0908	Incorrect, Inadequate or Imprecise Result or Readings	16/10/2023	16/10/2023
							F1905	Device Revision or Replacement	A0908	Incorrect, Inadequate or Imprecise Result or Readings	16/10/2023	16/10/2023
					CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A0908	Incorrect, Inadequate or Imprecise Result or Readings	16/10/2023	16/10/2023
							F08	Hospitalization or Prolonged Hospitalization	A0908	Incorrect, Inadequate or Imprecise Result or Readings	16/10/2023	16/10/2023
							F1905	Device Revision or Replacement	A0908	Incorrect, Inadequate or Imprecise Result or Readings	16/10/2023	16/10/2023
					CT233	Surgical	E2403	No Clinical Signs, Symptoms or Conditions	A0908	Incorrect, Inadequate or Imprecise Result or Readings	16/10/2023	16/10/2023
							F08	Hospitalization or Prolonged Hospitalization	A0908	Incorrect, Inadequate or Imprecise Result or Readings	16/10/2023	16/10/2023
							F1905	Device Revision or Replacement	A0908	Incorrect, Inadequate or Imprecise Result or Readings	16/10/2023	16/10/2023
					CT2406	Implantable	E2403	No Clinical Signs, Symptoms or Conditions	A0908	Incorrect, Inadequate or Imprecise Result or Readings	16/10/2023	16/10/2023
							F08	Hospitalization or Prolonged Hospitalization	A0908	Incorrect, Inadequate or Imprecise Result or Readings	16/10/2023	16/10/2023
							F1905	Device Revision or Replacement	A0908	Incorrect, Inadequate or Imprecise Result or Readings	16/10/2023	16/10/2023
					CT2407	Active implantable	E2403	No Clinical Signs, Symptoms or Conditions	A0908	Incorrect, Inadequate or Imprecise Result or Readings	16/10/2023	16/10/2023
							F08	Hospitalization or Prolonged Hospitalization	A0908	Incorrect, Inadequate or Imprecise Result or Readings	16/10/2023	16/10/2023
							F1905	Device Revision or Replacement	A0908	Incorrect, Inadequate or Imprecise Result or Readings	16/10/2023	16/10/2023
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A0908	Incorrect, Inadequate or Imprecise Result or Readings	16/10/2023	16/10/2023
							F08	Hospitalization or Prolonged Hospitalization	A0908	Incorrect, Inadequate or Imprecise Result or Readings	16/10/2023	16/10/2023
							F1905	Device Revision or Replacement	A0908	Incorrect, Inadequate or Imprecise Result or Readings	16/10/2023	16/10/2023
					CT2727	Active implant devices/ancillaries	E2403	No Clinical Signs, Symptoms or Conditions	A0908	Incorrect, Inadequate or Imprecise Result or Readings	16/10/2023	16/10/2023
							F08	Hospitalization or Prolonged Hospitalization	A0908	Incorrect, Inadequate or Imprecise Result or Readings	16/10/2023	16/10/2023
							F1905	Device Revision or Replacement	A0908	Incorrect, Inadequate or Imprecise Result or Readings	16/10/2023	16/10/2023
					CT2750	Implantable cardiac electrical stimulators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A0908	Incorrect, Inadequate or Imprecise Result or Readings	16/10/2023	16/10/2023
							F08	Hospitalization or Prolonged Hospitalization	A0908	Incorrect, Inadequate or Imprecise Result or Readings	16/10/2023	16/10/2023
							F1905	Device Revision or Replacement	A0908	Incorrect, Inadequate or Imprecise Result or Readings	16/10/2023	16/10/2023
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A0908	Incorrect, Inadequate or Imprecise Result or Readings	16/10/2023	16/10/2023
							F08	Hospitalization or Prolonged Hospitalization	A0908	Incorrect, Inadequate or Imprecise Result or Readings	16/10/2023	16/10/2023
							F1905	Device Revision or Replacement	A0908	Incorrect, Inadequate or Imprecise Result or Readings	16/10/2023	16/10/2023
					CT2985	Active	E2403	No Clinical Signs, Symptoms or Conditions	A0908	Incorrect, Inadequate or Imprecise Result or Readings	16/10/2023	16/10/2023
							F08	Hospitalization or Prolonged Hospitalization	A0908	Incorrect, Inadequate or Imprecise Result or Readings	16/10/2023	16/10/2023
							F1905	Device Revision or Replacement	A0908	Incorrect, Inadequate or Imprecise Result or Readings	16/10/2023	16/10/2023
					CT3001	Implantable metal	E2403	No Clinical Signs, Symptoms or Conditions	A0908	Incorrect, Inadequate or Imprecise Result or Readings	16/10/2023	16/10/2023
							F08	Hospitalization or Prolonged Hospitalization	A0908	Incorrect, Inadequate or Imprecise Result or Readings	16/10/2023	16/10/2023
							F1905	Device Revision or Replacement	A0908	Incorrect, Inadequate or Imprecise Result or Readings	16/10/2023	16/10/2023
					CT3002	Implantable synthetic polymer	E2403	No Clinical Signs, Symptoms or Conditions	A0908	Incorrect, Inadequate or Imprecise Result or Readings	16/10/2023	16/10/2023
							F08	Hospitalization or Prolonged Hospitalization	A0908	Incorrect, Inadequate or Imprecise Result or Readings	16/10/2023	16/10/2023
							F1905	Device Revision or Replacement	A0908	Incorrect, Inadequate or Imprecise Result or Readings	16/10/2023	16/10/2023

					CT3006	Implantable inorganic material	E2403	No Clinical Signs, Symptoms or Conditions	A0908	Incorrect, Inadequate or Imprecise Result or Readings	16/10/2023	16/10/2023
							F08	Hospitalization or Prolonged Hospitalization	A0908	Incorrect, Inadequate or Imprecise Result or Readings	16/10/2023	16/10/2023
							F1905	Device Revision or Replacement	A0908	Incorrect, Inadequate or Imprecise Result or Readings	16/10/2023	16/10/2023
					CT314	Electrically powered	E2403	No Clinical Signs, Symptoms or Conditions	A0908	Incorrect, Inadequate or Imprecise Result or Readings	16/10/2023	16/10/2023
							F08	Hospitalization or Prolonged Hospitalization	A0908	Incorrect, Inadequate or Imprecise Result or Readings	16/10/2023	16/10/2023
							F1905	Device Revision or Replacement	A0908	Incorrect, Inadequate or Imprecise Result or Readings	16/10/2023	16/10/2023
					CT321	Long-term surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A0908	Incorrect, Inadequate or Imprecise Result or Readings	16/10/2023	16/10/2023
							F08	Hospitalization or Prolonged Hospitalization	A0908	Incorrect, Inadequate or Imprecise Result or Readings	16/10/2023	16/10/2023
							F1905	Device Revision or Replacement	A0908	Incorrect, Inadequate or Imprecise Result or Readings	16/10/2023	16/10/2023
					CT449	Generators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A0908	Incorrect, Inadequate or Imprecise Result or Readings	16/10/2023	16/10/2023
							F08	Hospitalization or Prolonged Hospitalization	A0908	Incorrect, Inadequate or Imprecise Result or Readings	16/10/2023	16/10/2023
							F1905	Device Revision or Replacement	A0908	Incorrect, Inadequate or Imprecise Result or Readings	16/10/2023	16/10/2023
					CT981	Single use	E2403	No Clinical Signs, Symptoms or Conditions	A0908	Incorrect, Inadequate or Imprecise Result or Readings	16/10/2023	16/10/2023
							F08	Hospitalization or Prolonged Hospitalization	A0908	Incorrect, Inadequate or Imprecise Result or Readings	16/10/2023	16/10/2023
							F1905	Device Revision or Replacement	A0908	Incorrect, Inadequate or Imprecise Result or Readings	16/10/2023	16/10/2023
					CT983	Surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A0908	Incorrect, Inadequate or Imprecise Result or Readings	16/10/2023	16/10/2023
							F08	Hospitalization or Prolonged Hospitalization	A0908	Incorrect, Inadequate or Imprecise Result or Readings	16/10/2023	16/10/2023
							F1905	Device Revision or Replacement	A0908	Incorrect, Inadequate or Imprecise Result or Readings	16/10/2023	16/10/2023
14246	2023	All other reportable incidents	47263	Cardiac resynchronization therapy implantable pacemaker	CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A0908	Incorrect, Inadequate or Imprecise Result or Readings	17/07/2023	17/07/2023
							F08	Hospitalization or Prolonged Hospitalization	A0908	Incorrect, Inadequate or Imprecise Result or Readings	17/07/2023	17/07/2023
							F1905	Device Revision or Replacement	A0908	Incorrect, Inadequate or Imprecise Result or Readings	17/07/2023	17/07/2023
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A0908	Incorrect, Inadequate or Imprecise Result or Readings	17/07/2023	17/07/2023
							F08	Hospitalization or Prolonged Hospitalization	A0908	Incorrect, Inadequate or Imprecise Result or Readings	17/07/2023	17/07/2023
							F1905	Device Revision or Replacement	A0908	Incorrect, Inadequate or Imprecise Result or Readings	17/07/2023	17/07/2023
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A0908	Incorrect, Inadequate or Imprecise Result or Readings	17/07/2023	17/07/2023
							F08	Hospitalization or Prolonged Hospitalization	A0908	Incorrect, Inadequate or Imprecise Result or Readings	17/07/2023	17/07/2023
							F1905	Device Revision or Replacement	A0908	Incorrect, Inadequate or Imprecise Result or Readings	17/07/2023	17/07/2023
14247	2023	All other reportable incidents	36241	Pacemaker/defibrillator lead	CT1684	Cardiac pacemakers and associated devices	E2104	Electric Shock	A070909	Oversensing		
									A071302	Inappropriate/Inadequate Shock/Stimulation		
									A090801	Signal Artifact/Noise		
									A1301	Failure to Read Input Signal		
							F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A070909	Oversensing		
									A071302	Inappropriate/Inadequate Shock/Stimulation		
									A090801	Signal Artifact/Noise		
									A1301	Failure to Read Input Signal		
							F2203	Imaging Required	A070909	Oversensing		
									A071302	Inappropriate/Inadequate Shock/Stimulation		
									A090801	Signal Artifact/Noise		
									A1301	Failure to Read Input Signal		
14248	2023	All other reportable incidents	33500	Pacing/defibrillation lead repair kit	CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A051201	Device Dislodged or Dislocated		
									A070909	Oversensing		
									A090801	Signal Artifact/Noise		
							F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A051201	Device Dislodged or Dislocated		
									A070909	Oversensing		
									A090801	Signal Artifact/Noise		
14249	2023	All other reportable incidents	47265	Dual-chamber implantable pacemaker, rate-responsive	CT141	Electrodes/Leads and associated devices	E011903	Syncope/Fainting	A0722	Impedance Problem		
									A072201	High Impedance		
							F08	Hospitalization or Prolonged Hospitalization	A0722	Impedance Problem		
									A072201	High Impedance		
					CT145	Cardiovascular devices	E011903	Syncope/Fainting	A0722	Impedance Problem		
									A072201	High Impedance		
							F08	Hospitalization or Prolonged Hospitalization	A0722	Impedance Problem		
									A072201	High Impedance		
					CT1684	Cardiac pacemakers and associated devices	E011903	Syncope/Fainting	A0722	Impedance Problem		
									A072201	High Impedance		
							F08	Hospitalization or Prolonged Hospitalization	A0722	Impedance Problem		
									A072201	High Impedance		
					CT1685	Cardiac pacemakers	E011903	Syncope/Fainting	A0722	Impedance Problem		
									A072201	High Impedance		
							F08	Hospitalization or Prolonged Hospitalization	A0722	Impedance Problem		
									A072201	High Impedance		
					CT233	Surgical	E011903	Syncope/Fainting	A0722	Impedance Problem		
									A072201	High Impedance		
							F08	Hospitalization or Prolonged Hospitalization	A0722	Impedance Problem		
									A072201	High Impedance		
					CT2406	Implantable	E011903	Syncope/Fainting	A0722	Impedance Problem		
									A072201	High Impedance		

							F08	Hospitalization or Prolonged Hospitalization	A0722	Impedance Problem		
									A072201	High Impedance		
					CT2407	Active implantable	E011903	Syncope/Fainting	A0722	Impedance Problem		
									A072201	High Impedance		
							F08	Hospitalization or Prolonged Hospitalization	A0722	Impedance Problem		
									A072201	High Impedance		
					CT257	Implantable cardiac pacemakers	E011903	Syncope/Fainting	A0722	Impedance Problem		
									A072201	High Impedance		
							F08	Hospitalization or Prolonged Hospitalization	A0722	Impedance Problem		
									A072201	High Impedance		
					CT2727	Active implant devices/ancillaries	E011903	Syncope/Fainting	A0722	Impedance Problem		
									A072201	High Impedance		
							F08	Hospitalization or Prolonged Hospitalization	A0722	Impedance Problem		
									A072201	High Impedance		
					CT2750	Implantable cardiac electrical stimulators and associated devices	E011903	Syncope/Fainting	A0722	Impedance Problem		
									A072201	High Impedance		
							F08	Hospitalization or Prolonged Hospitalization	A0722	Impedance Problem		
									A072201	High Impedance		
					CT2866	Implantable cardiac electrical stimulators	E011903	Syncope/Fainting	A0722	Impedance Problem		
									A072201	High Impedance		
							F08	Hospitalization or Prolonged Hospitalization	A0722	Impedance Problem		
									A072201	High Impedance		
					CT2985	Active	E011903	Syncope/Fainting	A0722	Impedance Problem		
									A072201	High Impedance		
							F08	Hospitalization or Prolonged Hospitalization	A0722	Impedance Problem		
									A072201	High Impedance		
					CT3001	Implantable metal	E011903	Syncope/Fainting	A0722	Impedance Problem		
									A072201	High Impedance		
							F08	Hospitalization or Prolonged Hospitalization	A0722	Impedance Problem		
									A072201	High Impedance		
					CT3002	Implantable synthetic polymer	E011903	Syncope/Fainting	A0722	Impedance Problem		
									A072201	High Impedance		
							F08	Hospitalization or Prolonged Hospitalization	A0722	Impedance Problem		
									A072201	High Impedance		
					CT3006	Implantable inorganic material	E011903	Syncope/Fainting	A0722	Impedance Problem		
									A072201	High Impedance		
							F08	Hospitalization or Prolonged Hospitalization	A0722	Impedance Problem		
									A072201	High Impedance		
					CT314	Electrically powered	E011903	Syncope/Fainting	A0722	Impedance Problem		
									A072201	High Impedance		
							F08	Hospitalization or Prolonged Hospitalization	A0722	Impedance Problem		
									A072201	High Impedance		
					CT321	Long-term surgical invasive	E011903	Syncope/Fainting	A0722	Impedance Problem		
									A072201	High Impedance		
							F08	Hospitalization or Prolonged Hospitalization	A0722	Impedance Problem		
									A072201	High Impedance		
					CT449	Generators and associated devices	E011903	Syncope/Fainting	A0722	Impedance Problem		
									A072201	High Impedance		
							F08	Hospitalization or Prolonged Hospitalization	A0722	Impedance Problem		
									A072201	High Impedance		
					CT981	Single use	E011903	Syncope/Fainting	A0722	Impedance Problem		
									A072201	High Impedance		
							F08	Hospitalization or Prolonged Hospitalization	A0722	Impedance Problem		
									A072201	High Impedance		
					CT983	Surgical invasive	E011903	Syncope/Fainting	A0722	Impedance Problem		
									A072201	High Impedance		
							F08	Hospitalization or Prolonged Hospitalization	A0722	Impedance Problem		
									A072201	High Impedance		
14250	2023	All other reportable incidents	47265	Dual-chamber implantable pacemaker, rate-responsive	CT141	Electrodes/Leads and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070909	Oversensing		
									A090801	Signal Artifact/Noise		
							F15	Recognized Device or Procedural Complication	A070909	Oversensing		
									A090801	Signal Artifact/Noise		
					CT145	Cardiovascular devices	E2403	No Clinical Signs, Symptoms or Conditions	A070909	Oversensing		
									A090801	Signal Artifact/Noise		
							F15	Recognized Device or Procedural Complication	A070909	Oversensing		
									A090801	Signal Artifact/Noise		
					CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070909	Oversensing		
									A090801	Signal Artifact/Noise		
							F15	Recognized Device or Procedural Complication	A070909	Oversensing		
									A090801	Signal Artifact/Noise		
					CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A070909	Oversensing		
									A090801	Signal Artifact/Noise		
							F15	Recognized Device or Procedural Complication	A070909	Oversensing		
									A090801	Signal Artifact/Noise		
					CT233	Surgical	E2403	No Clinical Signs, Symptoms or Conditions	A070909	Oversensing		
									A090801	Signal Artifact/Noise		
							F15	Recognized Device or Procedural Complication	A070909	Oversensing		
									A090801	Signal Artifact/Noise		
					CT2406	Implantable	E2403	No Clinical Signs, Symptoms or Conditions	A070909	Oversensing		
									A090801	Signal Artifact/Noise		
							F15	Recognized Device or Procedural Complication	A070909	Oversensing		
									A090801	Signal Artifact/Noise		
					CT2407	Active implantable	E2403	No Clinical Signs, Symptoms or Conditions	A070909	Oversensing		
									A090801	Signal Artifact/Noise		
							F15	Recognized Device or Procedural Complication	A070909	Oversensing		
									A090801	Signal Artifact/Noise		
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A070909	Oversensing		
									A090801	Signal Artifact/Noise		

							F15	Recognized Device or Procedural Complication	A070909	Oversensing		
									A090801	Signal Artifact/Noise		
					CT2727	Active implant devices/ancillaries	E2403	No Clinical Signs, Symptoms or Conditions	A070909	Oversensing		
									A090801	Signal Artifact/Noise		
							F15	Recognized Device or Procedural Complication	A070909	Oversensing		
									A090801	Signal Artifact/Noise		
					CT2750	Implantable cardiac electrical stimulators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070909	Oversensing		
									A090801	Signal Artifact/Noise		
							F15	Recognized Device or Procedural Complication	A070909	Oversensing		
									A090801	Signal Artifact/Noise		
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A070909	Oversensing		
									A090801	Signal Artifact/Noise		
							F15	Recognized Device or Procedural Complication	A070909	Oversensing		
									A090801	Signal Artifact/Noise		
					CT2985	Active	E2403	No Clinical Signs, Symptoms or Conditions	A070909	Oversensing		
									A090801	Signal Artifact/Noise		
							F15	Recognized Device or Procedural Complication	A070909	Oversensing		
									A090801	Signal Artifact/Noise		
					CT3001	Implantable metal	E2403	No Clinical Signs, Symptoms or Conditions	A070909	Oversensing		
									A090801	Signal Artifact/Noise		
							F15	Recognized Device or Procedural Complication	A070909	Oversensing		
									A090801	Signal Artifact/Noise		
					CT3002	Implantable synthetic polymer	E2403	No Clinical Signs, Symptoms or Conditions	A070909	Oversensing		
									A090801	Signal Artifact/Noise		
							F15	Recognized Device or Procedural Complication	A070909	Oversensing		
									A090801	Signal Artifact/Noise		
					CT3006	Implantable inorganic material	E2403	No Clinical Signs, Symptoms or Conditions	A070909	Oversensing		
									A090801	Signal Artifact/Noise		
							F15	Recognized Device or Procedural Complication	A070909	Oversensing		
									A090801	Signal Artifact/Noise		
					CT314	Electrically powered	E2403	No Clinical Signs, Symptoms or Conditions	A070909	Oversensing		
									A090801	Signal Artifact/Noise		
							F15	Recognized Device or Procedural Complication	A070909	Oversensing		
									A090801	Signal Artifact/Noise		
					CT321	Long-term surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A070909	Oversensing		
									A090801	Signal Artifact/Noise		
							F15	Recognized Device or Procedural Complication	A070909	Oversensing		
									A090801	Signal Artifact/Noise		
					CT449	Generators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070909	Oversensing		
									A090801	Signal Artifact/Noise		
							F15	Recognized Device or Procedural Complication	A070909	Oversensing		
									A090801	Signal Artifact/Noise		
					CT981	Single use	E2403	No Clinical Signs, Symptoms or Conditions	A070909	Oversensing		
									A090801	Signal Artifact/Noise		
							F15	Recognized Device or Procedural Complication	A070909	Oversensing		
									A090801	Signal Artifact/Noise		
					CT983	Surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A070909	Oversensing		
									A090801	Signal Artifact/Noise		
							F15	Recognized Device or Procedural Complication	A070909	Oversensing		
									A090801	Signal Artifact/Noise		
14251	2023	All other reportable incidents	47265	Dual-chamber implantable pacemaker, rate-responsive	CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	23/08/2023	23/08/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	23/08/2023	23/08/2023
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	23/08/2023	23/08/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	23/08/2023	23/08/2023
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	23/08/2023	23/08/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	23/08/2023	23/08/2023
14252	2023	All other reportable incidents	17882	Physiologic-monitoring defibrillation system	CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070803	Failure to Power Up		
							F27	Problem Identified During Non-Clinical Procedure	A070803	Failure to Power Up		
14253	2023	All other reportable incidents	17882	Physiologic-monitoring defibrillation system	CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A0405	Degraded		
									A0719	Unexpected Shutdown		
							F27	Problem Identified During Non-Clinical Procedure	A0405	Degraded		
									A0719	Unexpected Shutdown		
14254	2023	All other reportable incidents	47265	Dual-chamber implantable pacemaker, rate-responsive	CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	03/09/2023	03/09/2023
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	03/09/2023	03/09/2023
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	03/09/2023	03/09/2023
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	03/09/2023	03/09/2023
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	03/09/2023	03/09/2023
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	03/09/2023	03/09/2023
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	03/09/2023	03/09/2023
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	03/09/2023	03/09/2023
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	03/09/2023	03/09/2023
14255	2023	All other reportable incidents	35223	Endocardial/interventricular septal pacing lead	CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A0401	Break		
									A070102	High Capture Threshold		
									A070901	Decreased Sensitivity		
									A070909	Oversensing		
									A0722	Impedance Problem		
									A090801	Signal Artifact/Noise		
							F26	No Health Consequences or Impact	A0401	Break		
									A070102	High Capture Threshold		
									A070901	Decreased Sensitivity		
									A070909	Oversensing		
									A0722	Impedance Problem		
									A090801	Signal Artifact/Noise		
14256	2023		47265	Dual-chamber implantable pacemaker, rate-responsive	CT141	Electrodes/Leads and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	14/09/2023	14/09/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	14/09/2023	14/09/2023
					CT145	Cardiovascular devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	14/09/2023	14/09/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	14/09/2023	14/09/2023
					CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	14/09/2023	14/09/2023

							F1901	Additional Surgery	A24	No Apparent Device Problem	14/09/2023	14/09/2023
					CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	14/09/2023	14/09/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	14/09/2023	14/09/2023
					CT233	Surgical	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	14/09/2023	14/09/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	14/09/2023	14/09/2023
					CT2406	Implantable	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	14/09/2023	14/09/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	14/09/2023	14/09/2023
					CT2407	Active implantable	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	14/09/2023	14/09/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	14/09/2023	14/09/2023
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	14/09/2023	14/09/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	14/09/2023	14/09/2023
					CT2727	Active implant devices/ancillaries	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	14/09/2023	14/09/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	14/09/2023	14/09/2023
					CT2750	Implantable cardiac electrical stimulators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	14/09/2023	14/09/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	14/09/2023	14/09/2023
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	14/09/2023	14/09/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	14/09/2023	14/09/2023
					CT2985	Active	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	14/09/2023	14/09/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	14/09/2023	14/09/2023
					CT3001	Implantable metal	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	14/09/2023	14/09/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	14/09/2023	14/09/2023
					CT3002	Implantable synthetic polymer	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	14/09/2023	14/09/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	14/09/2023	14/09/2023
					CT3006	Implantable inorganic material	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	14/09/2023	14/09/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	14/09/2023	14/09/2023
					CT314	Electrically powered	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	14/09/2023	14/09/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	14/09/2023	14/09/2023
					CT321	Long-term surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	14/09/2023	14/09/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	14/09/2023	14/09/2023
					CT449	Generators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	14/09/2023	14/09/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	14/09/2023	14/09/2023
					CT981	Single use	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	14/09/2023	14/09/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	14/09/2023	14/09/2023
					CT983	Surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	14/09/2023	14/09/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	14/09/2023	14/09/2023
14257	2023	All other reportable incidents	47265	Dual-chamber implantable pacemaker, rate-responsive	CT1685	Cardiac pacemakers	E0601	Arrhythmia	A070909	Oversensing	05/07/2023	05/07/2023
									A071102	Failure to Interrogate	05/07/2023	05/07/2023
									A071204	No Pacing	05/07/2023	05/07/2023
									A0723	Unexpected Mode Switch	05/07/2023	05/07/2023
							F1901	Additional Surgery	A070909	Oversensing	05/07/2023	05/07/2023
									A071102	Failure to Interrogate	05/07/2023	05/07/2023
									A071204	No Pacing	05/07/2023	05/07/2023
									A0723	Unexpected Mode Switch	05/07/2023	05/07/2023
					CT257	Implantable cardiac pacemakers	E0601	Arrhythmia	A070909	Oversensing	05/07/2023	05/07/2023
									A071102	Failure to Interrogate	05/07/2023	05/07/2023
									A071204	No Pacing	05/07/2023	05/07/2023
									A0723	Unexpected Mode Switch	05/07/2023	05/07/2023
							F1901	Additional Surgery	A070909	Oversensing	05/07/2023	05/07/2023
									A071102	Failure to Interrogate	05/07/2023	05/07/2023
									A071204	No Pacing	05/07/2023	05/07/2023
									A0723	Unexpected Mode Switch	05/07/2023	05/07/2023
					CT2866	Implantable cardiac electrical stimulators	E0601	Arrhythmia	A070909	Oversensing	05/07/2023	05/07/2023
									A071102	Failure to Interrogate	05/07/2023	05/07/2023
									A071204	No Pacing	05/07/2023	05/07/2023
									A0723	Unexpected Mode Switch	05/07/2023	05/07/2023
							F1901	Additional Surgery	A070909	Oversensing	05/07/2023	05/07/2023
									A071102	Failure to Interrogate	05/07/2023	05/07/2023
									A071204	No Pacing	05/07/2023	05/07/2023
									A0723	Unexpected Mode Switch	05/07/2023	05/07/2023
14258	2023	All other reportable incidents	47265	Dual-chamber implantable pacemaker, rate-responsive	CT1685	Cardiac pacemakers	E0601	Arrhythmia	A070504	Premature Discharge of Battery	11/07/2023	11/07/2023
									A071102	Failure to Interrogate	11/07/2023	11/07/2023
									A071204	No Pacing	11/07/2023	11/07/2023
							E2312	Fatigue	A070504	Premature Discharge of Battery	11/07/2023	11/07/2023
									A071102	Failure to Interrogate	11/07/2023	11/07/2023
									A071204	No Pacing	11/07/2023	11/07/2023
							F1901	Additional Surgery	A070504	Premature Discharge of Battery	11/07/2023	11/07/2023
									A071102	Failure to Interrogate	11/07/2023	11/07/2023
									A071204	No Pacing	11/07/2023	11/07/2023
					CT257	Implantable cardiac pacemakers	E0601	Arrhythmia	A070504	Premature Discharge of Battery	11/07/2023	11/07/2023
									A071102	Failure to Interrogate	11/07/2023	11/07/2023
									A071204	No Pacing	11/07/2023	11/07/2023
							E2312	Fatigue	A070504	Premature Discharge of Battery	11/07/2023	11/07/2023
									A071102	Failure to Interrogate	11/07/2023	11/07/2023
									A071204	No Pacing	11/07/2023	11/07/2023
							F1901	Additional Surgery	A070504	Premature Discharge of Battery	11/07/2023	11/07/2023
									A071102	Failure to Interrogate	11/07/2023	11/07/2023
									A071204	No Pacing	11/07/2023	11/07/2023
					CT2866	Implantable cardiac electrical stimulators	E0601	Arrhythmia	A070504	Premature Discharge of Battery	11/07/2023	11/07/2023
									A071102	Failure to Interrogate	11/07/2023	11/07/2023
									A071204	No Pacing	11/07/2023	11/07/2023
							E2312	Fatigue	A070504	Premature Discharge of Battery	11/07/2023	11/07/2023
									A071102	Failure to Interrogate	11/07/2023	11/07/2023
									A071204	No Pacing	11/07/2023	11/07/2023
							F1901	Additional Surgery	A070504	Premature Discharge of Battery	11/07/2023	11/07/2023
									A071102	Failure to Interrogate	11/07/2023	11/07/2023
									A071204	No Pacing	11/07/2023	11/07/2023
14259	2023	All other reportable incidents	47265	Dual-chamber implantable pacemaker, rate-responsive	CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A070909	Oversensing		
							F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A070909	Oversensing		

					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A070909	Oversensing		
							F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A070909	Oversensing		
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A070909	Oversensing		
							F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A070909	Oversensing		
14260	2023	All other reportable incidents	37265	Dual-chamber implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem	15/08/2023	15/08/2023
									A070504	Premature Discharge of Battery	15/08/2023	15/08/2023
							F08	Hospitalization or Prolonged Hospitalization	A0705	Battery Problem	15/08/2023	15/08/2023
									A070504	Premature Discharge of Battery	15/08/2023	15/08/2023
							F1905	Device Revision or Replacement	A0705	Battery Problem	15/08/2023	15/08/2023
									A070504	Premature Discharge of Battery	15/08/2023	15/08/2023
14261	2023	All other reportable incidents	36241	Pacemaker/defibrillator lead	CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A072201	High Impedance		
							F26	No Health Consequences or Impact	A072201	High Impedance		
14262	2023		47265	Dual-chamber implantable pacemaker, rate-responsive	CT141	Electrodes/Leads and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	14/10/2023	14/10/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	14/10/2023	14/10/2023
					CT145	Cardiovascular devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	14/10/2023	14/10/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	14/10/2023	14/10/2023
					CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	14/10/2023	14/10/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	14/10/2023	14/10/2023
					CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	14/10/2023	14/10/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	14/10/2023	14/10/2023
					CT233	Surgical	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	14/10/2023	14/10/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	14/10/2023	14/10/2023
					CT2406	Implantable	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	14/10/2023	14/10/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	14/10/2023	14/10/2023
					CT2407	Active implantable	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	14/10/2023	14/10/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	14/10/2023	14/10/2023
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	14/10/2023	14/10/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	14/10/2023	14/10/2023
					CT2727	Active implant devices/ancillaries	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	14/10/2023	14/10/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	14/10/2023	14/10/2023
					CT2750	Implantable cardiac electrical stimulators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	14/10/2023	14/10/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	14/10/2023	14/10/2023
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	14/10/2023	14/10/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	14/10/2023	14/10/2023
					CT2985	Active	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	14/10/2023	14/10/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	14/10/2023	14/10/2023
					CT3001	Implantable metal	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	14/10/2023	14/10/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	14/10/2023	14/10/2023
					CT3002	Implantable synthetic polymer	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	14/10/2023	14/10/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	14/10/2023	14/10/2023
					CT3006	Implantable inorganic material	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	14/10/2023	14/10/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	14/10/2023	14/10/2023
					CT314	Electrically powered	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	14/10/2023	14/10/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	14/10/2023	14/10/2023
					CT321	Long-term surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	14/10/2023	14/10/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	14/10/2023	14/10/2023
					CT449	Generators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	14/10/2023	14/10/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	14/10/2023	14/10/2023
					CT981	Single use	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	14/10/2023	14/10/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	14/10/2023	14/10/2023
					CT983	Surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	14/10/2023	14/10/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	14/10/2023	14/10/2023
14263	2023	All other reportable incidents	35852	Single-chamber implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A072201	High Impedance		
							F26	No Health Consequences or Impact	A072201	High Impedance		
14264	2023	All other reportable incidents	47265	Dual-chamber implantable pacemaker, rate-responsive	CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem	12/10/2023	12/10/2023
									A070504	Premature Discharge of Battery	12/10/2023	12/10/2023
							F08	Hospitalization or Prolonged Hospitalization	A0705	Battery Problem	12/10/2023	12/10/2023
									A070504	Premature Discharge of Battery	12/10/2023	12/10/2023
							F1905	Device Revision or Replacement	A0705	Battery Problem	12/10/2023	12/10/2023
									A070504	Premature Discharge of Battery	12/10/2023	12/10/2023
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem	12/10/2023	12/10/2023
									A070504	Premature Discharge of Battery	12/10/2023	12/10/2023
							F08	Hospitalization or Prolonged Hospitalization	A0705	Battery Problem	12/10/2023	12/10/2023
									A070504	Premature Discharge of Battery	12/10/2023	12/10/2023
							F1905	Device Revision or Replacement	A0705	Battery Problem	12/10/2023	12/10/2023
									A070504	Premature Discharge of Battery	12/10/2023	12/10/2023
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem	12/10/2023	12/10/2023
									A070504	Premature Discharge of Battery	12/10/2023	12/10/2023
							F08	Hospitalization or Prolonged Hospitalization	A0705	Battery Problem	12/10/2023	12/10/2023
									A070504	Premature Discharge of Battery	12/10/2023	12/10/2023
							F1905	Device Revision or Replacement	A0705	Battery Problem	12/10/2023	12/10/2023
									A070504	Premature Discharge of Battery	12/10/2023	12/10/2023
14265	2023	All other reportable incidents	35852	Single-chamber implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
14266	2023	All other reportable incidents	35223	Endocardial/interventricular septal pacing lead	CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A0401	Break		
									A070909	Oversensing		
									A0712	Pacing Problem		
							F1901	Additional Surgery	A0401	Break		
									A070909	Oversensing		
									A0712	Pacing Problem		
14267	2023		47265	Dual-chamber implantable pacemaker, rate-responsive	CT141	Electrodes/Leads and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	18/08/2023	18/08/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	18/08/2023	18/08/2023
					CT145	Cardiovascular devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	18/08/2023	18/08/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	18/08/2023	18/08/2023
					CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	18/08/2023	18/08/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	18/08/2023	18/08/2023

					CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	18/08/2023	18/08/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	18/08/2023	18/08/2023
					CT233	Surgical	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	18/08/2023	18/08/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	18/08/2023	18/08/2023
					CT2406	Implantable	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	18/08/2023	18/08/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	18/08/2023	18/08/2023
					CT2407	Active implantable	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	18/08/2023	18/08/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	18/08/2023	18/08/2023
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	18/08/2023	18/08/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	18/08/2023	18/08/2023
					CT2727	Active implant devices/ancillaries	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	18/08/2023	18/08/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	18/08/2023	18/08/2023
					CT2750	Implantable cardiac electrical stimulators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	18/08/2023	18/08/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	18/08/2023	18/08/2023
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	18/08/2023	18/08/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	18/08/2023	18/08/2023
					CT2985	Active	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	18/08/2023	18/08/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	18/08/2023	18/08/2023
					CT3001	Implantable metal	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	18/08/2023	18/08/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	18/08/2023	18/08/2023
					CT3002	Implantable synthetic polymer	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	18/08/2023	18/08/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	18/08/2023	18/08/2023
					CT3006	Implantable inorganic material	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	18/08/2023	18/08/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	18/08/2023	18/08/2023
					CT314	Electrically powered	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	18/08/2023	18/08/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	18/08/2023	18/08/2023
					CT321	Long-term surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	18/08/2023	18/08/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	18/08/2023	18/08/2023
					CT449	Generators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	18/08/2023	18/08/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	18/08/2023	18/08/2023
					CT981	Single use	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	18/08/2023	18/08/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	18/08/2023	18/08/2023
					CT983	Surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	18/08/2023	18/08/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	18/08/2023	18/08/2023
14268	2023		47265	Dual-chamber implantable pacemaker, rate-responsive	CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	13/10/2023	13/10/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	13/10/2023	13/10/2023
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	13/10/2023	13/10/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	13/10/2023	13/10/2023
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	13/10/2023	13/10/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	13/10/2023	13/10/2023
14269	2023		47265	Dual-chamber implantable pacemaker, rate-responsive	CT141	Electrodes/Leads and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	05/10/2023	05/10/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	05/10/2023	05/10/2023
					CT145	Cardiovascular devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	05/10/2023	05/10/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	05/10/2023	05/10/2023
					CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	05/10/2023	05/10/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	05/10/2023	05/10/2023
					CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	05/10/2023	05/10/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	05/10/2023	05/10/2023
					CT233	Surgical	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	05/10/2023	05/10/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	05/10/2023	05/10/2023
					CT2406	Implantable	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	05/10/2023	05/10/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	05/10/2023	05/10/2023
					CT2407	Active implantable	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	05/10/2023	05/10/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	05/10/2023	05/10/2023
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	05/10/2023	05/10/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	05/10/2023	05/10/2023
					CT2727	Active implant devices/ancillaries	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	05/10/2023	05/10/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	05/10/2023	05/10/2023
					CT2750	Implantable cardiac electrical stimulators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	05/10/2023	05/10/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	05/10/2023	05/10/2023
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	05/10/2023	05/10/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	05/10/2023	05/10/2023
					CT2985	Active	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	05/10/2023	05/10/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	05/10/2023	05/10/2023
					CT3001	Implantable metal	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	05/10/2023	05/10/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	05/10/2023	05/10/2023
					CT3002	Implantable synthetic polymer	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	05/10/2023	05/10/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	05/10/2023	05/10/2023
					CT3006	Implantable inorganic material	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	05/10/2023	05/10/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	05/10/2023	05/10/2023
					CT314	Electrically powered	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	05/10/2023	05/10/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	05/10/2023	05/10/2023
					CT321	Long-term surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	05/10/2023	05/10/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	05/10/2023	05/10/2023
					CT449	Generators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	05/10/2023	05/10/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	05/10/2023	05/10/2023
					CT981	Single use	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	05/10/2023	05/10/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	05/10/2023	05/10/2023
					CT983	Surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	05/10/2023	05/10/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	05/10/2023	05/10/2023
14270	2023	All other reportable incidents	17882	Physiologic-monitoring defibrillation system	CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A090811	Unable to Obtain Readings		
							F27	Problem Identified During Non-Clinical Procedure	A090811	Unable to Obtain Readings		
14271	2023		47265	Dual-chamber implantable pacemaker, rate-responsive	CT141	Electrodes/Leads and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	17/03/2023	17/03/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	17/03/2023	17/03/2023
					CT145	Cardiovascular devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	17/03/2023	17/03/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	17/03/2023	17/03/2023
					CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	17/03/2023	17/03/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	17/03/2023	17/03/2023

					CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	17/03/2023	17/03/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	17/03/2023	17/03/2023
					CT233	Surgical	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	17/03/2023	17/03/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	17/03/2023	17/03/2023
					CT2406	Implantable	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	17/03/2023	17/03/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	17/03/2023	17/03/2023
					CT2407	Active implantable	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	17/03/2023	17/03/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	17/03/2023	17/03/2023
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	17/03/2023	17/03/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	17/03/2023	17/03/2023
					CT2727	Active implant devices/ancillaries	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	17/03/2023	17/03/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	17/03/2023	17/03/2023
					CT2750	Implantable cardiac electrical stimulators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	17/03/2023	17/03/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	17/03/2023	17/03/2023
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	17/03/2023	17/03/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	17/03/2023	17/03/2023
					CT2985	Active	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	17/03/2023	17/03/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	17/03/2023	17/03/2023
					CT3001	Implantable metal	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	17/03/2023	17/03/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	17/03/2023	17/03/2023
					CT3002	Implantable synthetic polymer	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	17/03/2023	17/03/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	17/03/2023	17/03/2023
					CT3006	Implantable inorganic material	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	17/03/2023	17/03/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	17/03/2023	17/03/2023
					CT314	Electrically powered	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	17/03/2023	17/03/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	17/03/2023	17/03/2023
					CT321	Long-term surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	17/03/2023	17/03/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	17/03/2023	17/03/2023
					CT449	Generators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	17/03/2023	17/03/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	17/03/2023	17/03/2023
					CT981	Single use	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	17/03/2023	17/03/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	17/03/2023	17/03/2023
					CT983	Surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	17/03/2023	17/03/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	17/03/2023	17/03/2023
14272	2023		47265	Dual-chamber implantable pacemaker, rate-responsive	CT141	Electrodes/Leads and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	10/10/2023	10/10/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	10/10/2023	10/10/2023
					CT145	Cardiovascular devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	10/10/2023	10/10/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	10/10/2023	10/10/2023
					CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	10/10/2023	10/10/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	10/10/2023	10/10/2023
					CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	10/10/2023	10/10/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	10/10/2023	10/10/2023
					CT233	Surgical	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	10/10/2023	10/10/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	10/10/2023	10/10/2023
					CT2406	Implantable	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	10/10/2023	10/10/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	10/10/2023	10/10/2023
					CT2407	Active implantable	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	10/10/2023	10/10/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	10/10/2023	10/10/2023
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	10/10/2023	10/10/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	10/10/2023	10/10/2023
					CT2727	Active implant devices/ancillaries	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	10/10/2023	10/10/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	10/10/2023	10/10/2023
					CT2750	Implantable cardiac electrical stimulators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	10/10/2023	10/10/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	10/10/2023	10/10/2023
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	10/10/2023	10/10/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	10/10/2023	10/10/2023
					CT2985	Active	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	10/10/2023	10/10/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	10/10/2023	10/10/2023
					CT3001	Implantable metal	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	10/10/2023	10/10/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	10/10/2023	10/10/2023
					CT3002	Implantable synthetic polymer	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	10/10/2023	10/10/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	10/10/2023	10/10/2023
					CT3006	Implantable inorganic material	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	10/10/2023	10/10/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	10/10/2023	10/10/2023
					CT314	Electrically powered	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	10/10/2023	10/10/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	10/10/2023	10/10/2023
					CT321	Long-term surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	10/10/2023	10/10/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	10/10/2023	10/10/2023
					CT449	Generators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	10/10/2023	10/10/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	10/10/2023	10/10/2023
					CT981	Single use	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	10/10/2023	10/10/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	10/10/2023	10/10/2023
					CT983	Surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	10/10/2023	10/10/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	10/10/2023	10/10/2023
14273	2023		47265	Dual-chamber implantable pacemaker, rate-responsive	CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	14/10/2023	14/10/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	14/10/2023	14/10/2023
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	14/10/2023	14/10/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	14/10/2023	14/10/2023
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	14/10/2023	14/10/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	14/10/2023	14/10/2023
14274	2023	All other reportable incidents	47265	Dual-chamber implantable pacemaker, rate-responsive	CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem	17/11/2023	17/11/2023
									A070504	Premature Discharge of Battery	17/11/2023	17/11/2023
							F08	Hospitalization or Prolonged Hospitalization	A0705	Battery Problem	17/11/2023	17/11/2023
									A070504	Premature Discharge of Battery	17/11/2023	17/11/2023
							F1905	Device Revision or Replacement	A0705	Battery Problem	17/11/2023	17/11/2023
									A070504	Premature Discharge of Battery	17/11/2023	17/11/2023
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem	17/11/2023	17/11/2023
									A070504	Premature Discharge of Battery	17/11/2023	17/11/2023

							F08	Hospitalization or Prolonged Hospitalization	A0705	Battery Problem	17/11/2023	17/11/2023
									A070504	Premature Discharge of Battery	17/11/2023	17/11/2023
							F1905	Device Revision or Replacement	A0705	Battery Problem	17/11/2023	17/11/2023
									A070504	Premature Discharge of Battery	17/11/2023	17/11/2023
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem	17/11/2023	17/11/2023
									A070504	Premature Discharge of Battery	17/11/2023	17/11/2023
							F08	Hospitalization or Prolonged Hospitalization	A0705	Battery Problem	17/11/2023	17/11/2023
									A070504	Premature Discharge of Battery	17/11/2023	17/11/2023
							F1905	Device Revision or Replacement	A0705	Battery Problem	17/11/2023	17/11/2023
									A070504	Premature Discharge of Battery	17/11/2023	17/11/2023
14275	2023	All other reportable incidents	47265	Dual-chamber implantable pacemaker, rate-responsive	CT141	Electrodes/Leads and associated devices	E011902	Presyncope	A070101	Failure to Capture	10/10/2023	10/10/2023
									A0705	Battery Problem	10/10/2023	10/10/2023
									A0908	Incorrect, Inadequate or Imprecise Result or Readings	10/10/2023	10/10/2023
							E012207	Twitching	A070101	Failure to Capture	10/10/2023	10/10/2023
									A0705	Battery Problem	10/10/2023	10/10/2023
									A0908	Incorrect, Inadequate or Imprecise Result or Readings	10/10/2023	10/10/2023
							E060101	Asystole	A070101	Failure to Capture	10/10/2023	10/10/2023
									A0705	Battery Problem	10/10/2023	10/10/2023
									A0908	Incorrect, Inadequate or Imprecise Result or Readings	10/10/2023	10/10/2023
							F08	Hospitalization or Prolonged Hospitalization	A070101	Failure to Capture	10/10/2023	10/10/2023
									A0705	Battery Problem	10/10/2023	10/10/2023
									A0908	Incorrect, Inadequate or Imprecise Result or Readings	10/10/2023	10/10/2023
							F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A070101	Failure to Capture	10/10/2023	10/10/2023
									A0705	Battery Problem	10/10/2023	10/10/2023
									A0908	Incorrect, Inadequate or Imprecise Result or Readings	10/10/2023	10/10/2023
							F1905	Device Revision or Replacement	A070101	Failure to Capture	10/10/2023	10/10/2023
									A0705	Battery Problem	10/10/2023	10/10/2023
									A0908	Incorrect, Inadequate or Imprecise Result or Readings	10/10/2023	10/10/2023
					CT145	Cardiovascular devices	E011902	Presyncope	A070101	Failure to Capture	10/10/2023	10/10/2023
									A0705	Battery Problem	10/10/2023	10/10/2023
									A0908	Incorrect, Inadequate or Imprecise Result or Readings	10/10/2023	10/10/2023
							E012207	Twitching	A070101	Failure to Capture	10/10/2023	10/10/2023
									A0705	Battery Problem	10/10/2023	10/10/2023
									A0908	Incorrect, Inadequate or Imprecise Result or Readings	10/10/2023	10/10/2023
							E060101	Asystole	A070101	Failure to Capture	10/10/2023	10/10/2023
									A0705	Battery Problem	10/10/2023	10/10/2023
									A0908	Incorrect, Inadequate or Imprecise Result or Readings	10/10/2023	10/10/2023
							F08	Hospitalization or Prolonged Hospitalization	A070101	Failure to Capture	10/10/2023	10/10/2023
									A0705	Battery Problem	10/10/2023	10/10/2023
									A0908	Incorrect, Inadequate or Imprecise Result or Readings	10/10/2023	10/10/2023
							F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A070101	Failure to Capture	10/10/2023	10/10/2023
									A0705	Battery Problem	10/10/2023	10/10/2023
									A0908	Incorrect, Inadequate or Imprecise Result or Readings	10/10/2023	10/10/2023
							F1905	Device Revision or Replacement	A070101	Failure to Capture	10/10/2023	10/10/2023
									A0705	Battery Problem	10/10/2023	10/10/2023
									A0908	Incorrect, Inadequate or Imprecise Result or Readings	10/10/2023	10/10/2023
					CT1684	Cardiac pacemakers and associated devices	E011902	Presyncope	A070101	Failure to Capture	10/10/2023	10/10/2023
									A0705	Battery Problem	10/10/2023	10/10/2023
									A0908	Incorrect, Inadequate or Imprecise Result or Readings	10/10/2023	10/10/2023
							E012207	Twitching	A070101	Failure to Capture	10/10/2023	10/10/2023
									A0705	Battery Problem	10/10/2023	10/10/2023
									A0908	Incorrect, Inadequate or Imprecise Result or Readings	10/10/2023	10/10/2023
							E060101	Asystole	A070101	Failure to Capture	10/10/2023	10/10/2023
									A0705	Battery Problem	10/10/2023	10/10/2023
									A0908	Incorrect, Inadequate or Imprecise Result or Readings	10/10/2023	10/10/2023
							F08	Hospitalization or Prolonged Hospitalization	A070101	Failure to Capture	10/10/2023	10/10/2023
									A0705	Battery Problem	10/10/2023	10/10/2023
									A0908	Incorrect, Inadequate or Imprecise Result or Readings	10/10/2023	10/10/2023
							F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A070101	Failure to Capture	10/10/2023	10/10/2023
									A0705	Battery Problem	10/10/2023	10/10/2023
									A0908	Incorrect, Inadequate or Imprecise Result or Readings	10/10/2023	10/10/2023
							F1905	Device Revision or Replacement	A070101	Failure to Capture	10/10/2023	10/10/2023
									A0705	Battery Problem	10/10/2023	10/10/2023
									A0908	Incorrect, Inadequate or Imprecise Result or Readings	10/10/2023	10/10/2023
					CT1685	Cardiac pacemakers	E011902	Presyncope	A070101	Failure to Capture	10/10/2023	10/10/2023
									A0705	Battery Problem	10/10/2023	10/10/2023
									A0908	Incorrect, Inadequate or Imprecise Result or Readings	10/10/2023	10/10/2023
							E012207	Twitching	A070101	Failure to Capture	10/10/2023	10/10/2023
									A0705	Battery Problem	10/10/2023	10/10/2023
									A0908	Incorrect, Inadequate or Imprecise Result or Readings	10/10/2023	10/10/2023
							E060101	Asystole	A070101	Failure to Capture	10/10/2023	10/10/2023
									A0705	Battery Problem	10/10/2023	10/10/2023

								A0908	Incorrect, Inadequate or Imprecise Result or Readings	10/10/2023	10/10/2023
						F08	Hospitalization or Prolonged Hospitalization	A070101	Failure to Capture	10/10/2023	10/10/2023
								A0705	Battery Problem	10/10/2023	10/10/2023
								A0908	Incorrect, Inadequate or Imprecise Result or Readings	10/10/2023	10/10/2023
						F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A070101	Failure to Capture	10/10/2023	10/10/2023
								A0705	Battery Problem	10/10/2023	10/10/2023
								A0908	Incorrect, Inadequate or Imprecise Result or Readings	10/10/2023	10/10/2023
						F1905	Device Revision or Replacement	A070101	Failure to Capture	10/10/2023	10/10/2023
								A0705	Battery Problem	10/10/2023	10/10/2023
								A0908	Incorrect, Inadequate or Imprecise Result or Readings	10/10/2023	10/10/2023
					CT233	Surgical	E011902	Presyncope	A070101	Failure to Capture	10/10/2023	10/10/2023
									A0705	Battery Problem	10/10/2023	10/10/2023
									A0908	Incorrect, Inadequate or Imprecise Result or Readings	10/10/2023	10/10/2023
							E012207	Twitching	A070101	Failure to Capture	10/10/2023	10/10/2023
									A0705	Battery Problem	10/10/2023	10/10/2023
									A0908	Incorrect, Inadequate or Imprecise Result or Readings	10/10/2023	10/10/2023
							E060101	Asystole	A070101	Failure to Capture	10/10/2023	10/10/2023
									A0705	Battery Problem	10/10/2023	10/10/2023
									A0908	Incorrect, Inadequate or Imprecise Result or Readings	10/10/2023	10/10/2023
							F08	Hospitalization or Prolonged Hospitalization	A070101	Failure to Capture	10/10/2023	10/10/2023
									A0705	Battery Problem	10/10/2023	10/10/2023
									A0908	Incorrect, Inadequate or Imprecise Result or Readings	10/10/2023	10/10/2023
							F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A070101	Failure to Capture	10/10/2023	10/10/2023
									A0705	Battery Problem	10/10/2023	10/10/2023
									A0908	Incorrect, Inadequate or Imprecise Result or Readings	10/10/2023	10/10/2023
							F1905	Device Revision or Replacement	A070101	Failure to Capture	10/10/2023	10/10/2023
									A0705	Battery Problem	10/10/2023	10/10/2023
									A0908	Incorrect, Inadequate or Imprecise Result or Readings	10/10/2023	10/10/2023
					CT2406	Implantable	E011902	Presyncope	A070101	Failure to Capture	10/10/2023	10/10/2023
									A0705	Battery Problem	10/10/2023	10/10/2023
									A0908	Incorrect, Inadequate or Imprecise Result or Readings	10/10/2023	10/10/2023
							E012207	Twitching	A070101	Failure to Capture	10/10/2023	10/10/2023
									A0705	Battery Problem	10/10/2023	10/10/2023
									A0908	Incorrect, Inadequate or Imprecise Result or Readings	10/10/2023	10/10/2023
							E060101	Asystole	A070101	Failure to Capture	10/10/2023	10/10/2023
									A0705	Battery Problem	10/10/2023	10/10/2023
									A0908	Incorrect, Inadequate or Imprecise Result or Readings	10/10/2023	10/10/2023
							F08	Hospitalization or Prolonged Hospitalization	A070101	Failure to Capture	10/10/2023	10/10/2023
									A0705	Battery Problem	10/10/2023	10/10/2023
									A0908	Incorrect, Inadequate or Imprecise Result or Readings	10/10/2023	10/10/2023
							F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A070101	Failure to Capture	10/10/2023	10/10/2023
									A0705	Battery Problem	10/10/2023	10/10/2023
									A0908	Incorrect, Inadequate or Imprecise Result or Readings	10/10/2023	10/10/2023
							F1905	Device Revision or Replacement	A070101	Failure to Capture	10/10/2023	10/10/2023
									A0705	Battery Problem	10/10/2023	10/10/2023
									A0908	Incorrect, Inadequate or Imprecise Result or Readings	10/10/2023	10/10/2023
					CT2407	Active implantable	E011902	Presyncope	A070101	Failure to Capture	10/10/2023	10/10/2023
									A0705	Battery Problem	10/10/2023	10/10/2023
									A0908	Incorrect, Inadequate or Imprecise Result or Readings	10/10/2023	10/10/2023
							E012207	Twitching	A070101	Failure to Capture	10/10/2023	10/10/2023
									A0705	Battery Problem	10/10/2023	10/10/2023
									A0908	Incorrect, Inadequate or Imprecise Result or Readings	10/10/2023	10/10/2023
							E060101	Asystole	A070101	Failure to Capture	10/10/2023	10/10/2023
									A0705	Battery Problem	10/10/2023	10/10/2023
									A0908	Incorrect, Inadequate or Imprecise Result or Readings	10/10/2023	10/10/2023
							F08	Hospitalization or Prolonged Hospitalization	A070101	Failure to Capture	10/10/2023	10/10/2023
									A0705	Battery Problem	10/10/2023	10/10/2023
									A0908	Incorrect, Inadequate or Imprecise Result or Readings	10/10/2023	10/10/2023
							F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A070101	Failure to Capture	10/10/2023	10/10/2023
									A0705	Battery Problem	10/10/2023	10/10/2023
									A0908	Incorrect, Inadequate or Imprecise Result or Readings	10/10/2023	10/10/2023
							F1905	Device Revision or Replacement	A070101	Failure to Capture	10/10/2023	10/10/2023
									A0705	Battery Problem	10/10/2023	10/10/2023
									A0908	Incorrect, Inadequate or Imprecise Result or Readings	10/10/2023	10/10/2023
					CT257	Implantable cardiac pacemakers	E011902	Presyncope	A070101	Failure to Capture	10/10/2023	10/10/2023
									A0705	Battery Problem	10/10/2023	10/10/2023
									A0908	Incorrect, Inadequate or Imprecise Result or Readings	10/10/2023	10/10/2023
							E012207	Twitching	A070101	Failure to Capture	10/10/2023	10/10/2023
									A0705	Battery Problem	10/10/2023	10/10/2023

								A0908	Incorrect, Inadequate or Imprecise Result or Readings	10/10/2023	10/10/2023
						E060101	Asystole	A070101	Failure to Capture	10/10/2023	10/10/2023
								A0705	Battery Problem	10/10/2023	10/10/2023
								A0908	Incorrect, Inadequate or Imprecise Result or Readings	10/10/2023	10/10/2023
						F08	Hospitalization or Prolonged Hospitalization	A070101	Failure to Capture	10/10/2023	10/10/2023
								A0705	Battery Problem	10/10/2023	10/10/2023
								A0908	Incorrect, Inadequate or Imprecise Result or Readings	10/10/2023	10/10/2023
						F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A070101	Failure to Capture	10/10/2023	10/10/2023
								A0705	Battery Problem	10/10/2023	10/10/2023
								A0908	Incorrect, Inadequate or Imprecise Result or Readings	10/10/2023	10/10/2023
						F1905	Device Revision or Replacement	A070101	Failure to Capture	10/10/2023	10/10/2023
								A0705	Battery Problem	10/10/2023	10/10/2023
								A0908	Incorrect, Inadequate or Imprecise Result or Readings	10/10/2023	10/10/2023
					CT2727	Active implant devices/ancillaries	E011902	Presyncope	A070101	Failure to Capture	10/10/2023	10/10/2023
									A0705	Battery Problem	10/10/2023	10/10/2023
									A0908	Incorrect, Inadequate or Imprecise Result or Readings	10/10/2023	10/10/2023
							E012207	Twitching	A070101	Failure to Capture	10/10/2023	10/10/2023
									A0705	Battery Problem	10/10/2023	10/10/2023
									A0908	Incorrect, Inadequate or Imprecise Result or Readings	10/10/2023	10/10/2023
							E060101	Asystole	A070101	Failure to Capture	10/10/2023	10/10/2023
									A0705	Battery Problem	10/10/2023	10/10/2023
									A0908	Incorrect, Inadequate or Imprecise Result or Readings	10/10/2023	10/10/2023
							F08	Hospitalization or Prolonged Hospitalization	A070101	Failure to Capture	10/10/2023	10/10/2023
									A0705	Battery Problem	10/10/2023	10/10/2023
									A0908	Incorrect, Inadequate or Imprecise Result or Readings	10/10/2023	10/10/2023
							F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A070101	Failure to Capture	10/10/2023	10/10/2023
									A0705	Battery Problem	10/10/2023	10/10/2023
									A0908	Incorrect, Inadequate or Imprecise Result or Readings	10/10/2023	10/10/2023
							F1905	Device Revision or Replacement	A070101	Failure to Capture	10/10/2023	10/10/2023
									A0705	Battery Problem	10/10/2023	10/10/2023
									A0908	Incorrect, Inadequate or Imprecise Result or Readings	10/10/2023	10/10/2023
					CT2750	Implantable cardiac electrical stimulators and associated devices	E011902	Presyncope	A070101	Failure to Capture	10/10/2023	10/10/2023
									A0705	Battery Problem	10/10/2023	10/10/2023
									A0908	Incorrect, Inadequate or Imprecise Result or Readings	10/10/2023	10/10/2023
							E012207	Twitching	A070101	Failure to Capture	10/10/2023	10/10/2023
									A0705	Battery Problem	10/10/2023	10/10/2023
									A0908	Incorrect, Inadequate or Imprecise Result or Readings	10/10/2023	10/10/2023
							E060101	Asystole	A070101	Failure to Capture	10/10/2023	10/10/2023
									A0705	Battery Problem	10/10/2023	10/10/2023
									A0908	Incorrect, Inadequate or Imprecise Result or Readings	10/10/2023	10/10/2023
							F08	Hospitalization or Prolonged Hospitalization	A070101	Failure to Capture	10/10/2023	10/10/2023
									A0705	Battery Problem	10/10/2023	10/10/2023
									A0908	Incorrect, Inadequate or Imprecise Result or Readings	10/10/2023	10/10/2023
							F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A070101	Failure to Capture	10/10/2023	10/10/2023
									A0705	Battery Problem	10/10/2023	10/10/2023
									A0908	Incorrect, Inadequate or Imprecise Result or Readings	10/10/2023	10/10/2023
							F1905	Device Revision or Replacement	A070101	Failure to Capture	10/10/2023	10/10/2023
									A0705	Battery Problem	10/10/2023	10/10/2023
									A0908	Incorrect, Inadequate or Imprecise Result or Readings	10/10/2023	10/10/2023
					CT2866	Implantable cardiac electrical stimulators	E011902	Presyncope	A070101	Failure to Capture	10/10/2023	10/10/2023
									A0705	Battery Problem	10/10/2023	10/10/2023
									A0908	Incorrect, Inadequate or Imprecise Result or Readings	10/10/2023	10/10/2023
							E012207	Twitching	A070101	Failure to Capture	10/10/2023	10/10/2023
									A0705	Battery Problem	10/10/2023	10/10/2023
									A0908	Incorrect, Inadequate or Imprecise Result or Readings	10/10/2023	10/10/2023
							E060101	Asystole	A070101	Failure to Capture	10/10/2023	10/10/2023
									A0705	Battery Problem	10/10/2023	10/10/2023
									A0908	Incorrect, Inadequate or Imprecise Result or Readings	10/10/2023	10/10/2023
							F08	Hospitalization or Prolonged Hospitalization	A070101	Failure to Capture	10/10/2023	10/10/2023
									A0705	Battery Problem	10/10/2023	10/10/2023
									A0908	Incorrect, Inadequate or Imprecise Result or Readings	10/10/2023	10/10/2023
							F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A070101	Failure to Capture	10/10/2023	10/10/2023
									A0705	Battery Problem	10/10/2023	10/10/2023
									A0908	Incorrect, Inadequate or Imprecise Result or Readings	10/10/2023	10/10/2023
							F1905	Device Revision or Replacement	A070101	Failure to Capture	10/10/2023	10/10/2023
									A0705	Battery Problem	10/10/2023	10/10/2023
									A0908	Incorrect, Inadequate or Imprecise Result or Readings	10/10/2023	10/10/2023
					CT2985	Active	E011902	Presyncope	A070101	Failure to Capture	10/10/2023	10/10/2023
									A0705	Battery Problem	10/10/2023	10/10/2023

									A0908	Incorrect, Inadequate or Imprecise Result or Readings	10/10/2023	10/10/2023
							E012207	Twitching	A070101	Failure to Capture	10/10/2023	10/10/2023
									A0705	Battery Problem	10/10/2023	10/10/2023
									A0908	Incorrect, Inadequate or Imprecise Result or Readings	10/10/2023	10/10/2023
							E060101	Asystole	A070101	Failure to Capture	10/10/2023	10/10/2023
									A0705	Battery Problem	10/10/2023	10/10/2023
									A0908	Incorrect, Inadequate or Imprecise Result or Readings	10/10/2023	10/10/2023
							F08	Hospitalization or Prolonged Hospitalization	A070101	Failure to Capture	10/10/2023	10/10/2023
									A0705	Battery Problem	10/10/2023	10/10/2023
									A0908	Incorrect, Inadequate or Imprecise Result or Readings	10/10/2023	10/10/2023
							F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A070101	Failure to Capture	10/10/2023	10/10/2023
									A0705	Battery Problem	10/10/2023	10/10/2023
									A0908	Incorrect, Inadequate or Imprecise Result or Readings	10/10/2023	10/10/2023
							F1905	Device Revision or Replacement	A070101	Failure to Capture	10/10/2023	10/10/2023
									A0705	Battery Problem	10/10/2023	10/10/2023
									A0908	Incorrect, Inadequate or Imprecise Result or Readings	10/10/2023	10/10/2023
					CT3001	Implantable metal	E011902	Presyncope	A070101	Failure to Capture	10/10/2023	10/10/2023
									A0705	Battery Problem	10/10/2023	10/10/2023
									A0908	Incorrect, Inadequate or Imprecise Result or Readings	10/10/2023	10/10/2023
							E012207	Twitching	A070101	Failure to Capture	10/10/2023	10/10/2023
									A0705	Battery Problem	10/10/2023	10/10/2023
									A0908	Incorrect, Inadequate or Imprecise Result or Readings	10/10/2023	10/10/2023
							E060101	Asystole	A070101	Failure to Capture	10/10/2023	10/10/2023
									A0705	Battery Problem	10/10/2023	10/10/2023
									A0908	Incorrect, Inadequate or Imprecise Result or Readings	10/10/2023	10/10/2023
							F08	Hospitalization or Prolonged Hospitalization	A070101	Failure to Capture	10/10/2023	10/10/2023
									A0705	Battery Problem	10/10/2023	10/10/2023
									A0908	Incorrect, Inadequate or Imprecise Result or Readings	10/10/2023	10/10/2023
							F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A070101	Failure to Capture	10/10/2023	10/10/2023
									A0705	Battery Problem	10/10/2023	10/10/2023
									A0908	Incorrect, Inadequate or Imprecise Result or Readings	10/10/2023	10/10/2023
							F1905	Device Revision or Replacement	A070101	Failure to Capture	10/10/2023	10/10/2023
									A0705	Battery Problem	10/10/2023	10/10/2023
									A0908	Incorrect, Inadequate or Imprecise Result or Readings	10/10/2023	10/10/2023
					CT3002	Implantable synthetic polymer	E011902	Presyncope	A070101	Failure to Capture	10/10/2023	10/10/2023
									A0705	Battery Problem	10/10/2023	10/10/2023
									A0908	Incorrect, Inadequate or Imprecise Result or Readings	10/10/2023	10/10/2023
							E012207	Twitching	A070101	Failure to Capture	10/10/2023	10/10/2023
									A0705	Battery Problem	10/10/2023	10/10/2023
									A0908	Incorrect, Inadequate or Imprecise Result or Readings	10/10/2023	10/10/2023
							E060101	Asystole	A070101	Failure to Capture	10/10/2023	10/10/2023
									A0705	Battery Problem	10/10/2023	10/10/2023
									A0908	Incorrect, Inadequate or Imprecise Result or Readings	10/10/2023	10/10/2023
							F08	Hospitalization or Prolonged Hospitalization	A070101	Failure to Capture	10/10/2023	10/10/2023
									A0705	Battery Problem	10/10/2023	10/10/2023
									A0908	Incorrect, Inadequate or Imprecise Result or Readings	10/10/2023	10/10/2023
							F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A070101	Failure to Capture	10/10/2023	10/10/2023
									A0705	Battery Problem	10/10/2023	10/10/2023
									A0908	Incorrect, Inadequate or Imprecise Result or Readings	10/10/2023	10/10/2023
							F1905	Device Revision or Replacement	A070101	Failure to Capture	10/10/2023	10/10/2023
									A0705	Battery Problem	10/10/2023	10/10/2023
									A0908	Incorrect, Inadequate or Imprecise Result or Readings	10/10/2023	10/10/2023
					CT3006	Implantable inorganic material	E011902	Presyncope	A070101	Failure to Capture	10/10/2023	10/10/2023
									A0705	Battery Problem	10/10/2023	10/10/2023
									A0908	Incorrect, Inadequate or Imprecise Result or Readings	10/10/2023	10/10/2023
							E012207	Twitching	A070101	Failure to Capture	10/10/2023	10/10/2023
									A0705	Battery Problem	10/10/2023	10/10/2023
									A0908	Incorrect, Inadequate or Imprecise Result or Readings	10/10/2023	10/10/2023
							E060101	Asystole	A070101	Failure to Capture	10/10/2023	10/10/2023
									A0705	Battery Problem	10/10/2023	10/10/2023
									A0908	Incorrect, Inadequate or Imprecise Result or Readings	10/10/2023	10/10/2023
							F08	Hospitalization or Prolonged Hospitalization	A070101	Failure to Capture	10/10/2023	10/10/2023
									A0705	Battery Problem	10/10/2023	10/10/2023
									A0908	Incorrect, Inadequate or Imprecise Result or Readings	10/10/2023	10/10/2023
							F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A070101	Failure to Capture	10/10/2023	10/10/2023
									A0705	Battery Problem	10/10/2023	10/10/2023
									A0908	Incorrect, Inadequate or Imprecise Result or Readings	10/10/2023	10/10/2023
							F1905	Device Revision or Replacement	A070101	Failure to Capture	10/10/2023	10/10/2023
									A0705	Battery Problem	10/10/2023	10/10/2023

									A0908	Incorrect, Inadequate or Imprecise Result or Readings	10/10/2023	10/10/2023
					CT314	Electrically powered	E011902	Presyncope	A070101	Failure to Capture	10/10/2023	10/10/2023
									A0705	Battery Problem	10/10/2023	10/10/2023
									A0908	Incorrect, Inadequate or Imprecise Result or Readings	10/10/2023	10/10/2023
							E012207	Twitching	A070101	Failure to Capture	10/10/2023	10/10/2023
									A0705	Battery Problem	10/10/2023	10/10/2023
									A0908	Incorrect, Inadequate or Imprecise Result or Readings	10/10/2023	10/10/2023
							E060101	Asystole	A070101	Failure to Capture	10/10/2023	10/10/2023
									A0705	Battery Problem	10/10/2023	10/10/2023
									A0908	Incorrect, Inadequate or Imprecise Result or Readings	10/10/2023	10/10/2023
							F08	Hospitalization or Prolonged Hospitalization	A070101	Failure to Capture	10/10/2023	10/10/2023
									A0705	Battery Problem	10/10/2023	10/10/2023
									A0908	Incorrect, Inadequate or Imprecise Result or Readings	10/10/2023	10/10/2023
							F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A070101	Failure to Capture	10/10/2023	10/10/2023
									A0705	Battery Problem	10/10/2023	10/10/2023
									A0908	Incorrect, Inadequate or Imprecise Result or Readings	10/10/2023	10/10/2023
							F1905	Device Revision or Replacement	A070101	Failure to Capture	10/10/2023	10/10/2023
									A0705	Battery Problem	10/10/2023	10/10/2023
									A0908	Incorrect, Inadequate or Imprecise Result or Readings	10/10/2023	10/10/2023
					CT321	Long-term surgical invasive	E011902	Presyncope	A070101	Failure to Capture	10/10/2023	10/10/2023
									A0705	Battery Problem	10/10/2023	10/10/2023
									A0908	Incorrect, Inadequate or Imprecise Result or Readings	10/10/2023	10/10/2023
							E012207	Twitching	A070101	Failure to Capture	10/10/2023	10/10/2023
									A0705	Battery Problem	10/10/2023	10/10/2023
									A0908	Incorrect, Inadequate or Imprecise Result or Readings	10/10/2023	10/10/2023
							E060101	Asystole	A070101	Failure to Capture	10/10/2023	10/10/2023
									A0705	Battery Problem	10/10/2023	10/10/2023
									A0908	Incorrect, Inadequate or Imprecise Result or Readings	10/10/2023	10/10/2023
							F08	Hospitalization or Prolonged Hospitalization	A070101	Failure to Capture	10/10/2023	10/10/2023
									A0705	Battery Problem	10/10/2023	10/10/2023
									A0908	Incorrect, Inadequate or Imprecise Result or Readings	10/10/2023	10/10/2023
							F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A070101	Failure to Capture	10/10/2023	10/10/2023
									A0705	Battery Problem	10/10/2023	10/10/2023
									A0908	Incorrect, Inadequate or Imprecise Result or Readings	10/10/2023	10/10/2023
							F1905	Device Revision or Replacement	A070101	Failure to Capture	10/10/2023	10/10/2023
									A0705	Battery Problem	10/10/2023	10/10/2023
									A0908	Incorrect, Inadequate or Imprecise Result or Readings	10/10/2023	10/10/2023
					CT449	Generators and associated devices	E011902	Presyncope	A070101	Failure to Capture	10/10/2023	10/10/2023
									A0705	Battery Problem	10/10/2023	10/10/2023
									A0908	Incorrect, Inadequate or Imprecise Result or Readings	10/10/2023	10/10/2023
							E012207	Twitching	A070101	Failure to Capture	10/10/2023	10/10/2023
									A0705	Battery Problem	10/10/2023	10/10/2023
									A0908	Incorrect, Inadequate or Imprecise Result or Readings	10/10/2023	10/10/2023
							E060101	Asystole	A070101	Failure to Capture	10/10/2023	10/10/2023
									A0705	Battery Problem	10/10/2023	10/10/2023
									A0908	Incorrect, Inadequate or Imprecise Result or Readings	10/10/2023	10/10/2023
							F08	Hospitalization or Prolonged Hospitalization	A070101	Failure to Capture	10/10/2023	10/10/2023
									A0705	Battery Problem	10/10/2023	10/10/2023
									A0908	Incorrect, Inadequate or Imprecise Result or Readings	10/10/2023	10/10/2023
							F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A070101	Failure to Capture	10/10/2023	10/10/2023
									A0705	Battery Problem	10/10/2023	10/10/2023
									A0908	Incorrect, Inadequate or Imprecise Result or Readings	10/10/2023	10/10/2023
							F1905	Device Revision or Replacement	A070101	Failure to Capture	10/10/2023	10/10/2023
									A0705	Battery Problem	10/10/2023	10/10/2023
									A0908	Incorrect, Inadequate or Imprecise Result or Readings	10/10/2023	10/10/2023
					CT981	Single use	E011902	Presyncope	A070101	Failure to Capture	10/10/2023	10/10/2023
									A0705	Battery Problem	10/10/2023	10/10/2023
									A0908	Incorrect, Inadequate or Imprecise Result or Readings	10/10/2023	10/10/2023
							E012207	Twitching	A070101	Failure to Capture	10/10/2023	10/10/2023
									A0705	Battery Problem	10/10/2023	10/10/2023
									A0908	Incorrect, Inadequate or Imprecise Result or Readings	10/10/2023	10/10/2023
							E060101	Asystole	A070101	Failure to Capture	10/10/2023	10/10/2023
									A0705	Battery Problem	10/10/2023	10/10/2023
									A0908	Incorrect, Inadequate or Imprecise Result or Readings	10/10/2023	10/10/2023
							F08	Hospitalization or Prolonged Hospitalization	A070101	Failure to Capture	10/10/2023	10/10/2023
									A0705	Battery Problem	10/10/2023	10/10/2023
									A0908	Incorrect, Inadequate or Imprecise Result or Readings	10/10/2023	10/10/2023
							F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A070101	Failure to Capture	10/10/2023	10/10/2023
									A0705	Battery Problem	10/10/2023	10/10/2023

								A0908	Incorrect, Inadequate or Imprecise Result or Readings	10/10/2023	10/10/2023
						F1905	Device Revision or Replacement	A070101	Failure to Capture	10/10/2023	10/10/2023
								A0705	Battery Problem	10/10/2023	10/10/2023
								A0908	Incorrect, Inadequate or Imprecise Result or Readings	10/10/2023	10/10/2023
				CT983	Surgical invasive	E011902	Presyncope	A070101	Failure to Capture	10/10/2023	10/10/2023
								A0705	Battery Problem	10/10/2023	10/10/2023
								A0908	Incorrect, Inadequate or Imprecise Result or Readings	10/10/2023	10/10/2023
						E012207	Twitching	A070101	Failure to Capture	10/10/2023	10/10/2023
								A0705	Battery Problem	10/10/2023	10/10/2023
								A0908	Incorrect, Inadequate or Imprecise Result or Readings	10/10/2023	10/10/2023
						E060101	Asystole	A070101	Failure to Capture	10/10/2023	10/10/2023
								A0705	Battery Problem	10/10/2023	10/10/2023
								A0908	Incorrect, Inadequate or Imprecise Result or Readings	10/10/2023	10/10/2023
						F08	Hospitalization or Prolonged Hospitalization	A070101	Failure to Capture	10/10/2023	10/10/2023
								A0705	Battery Problem	10/10/2023	10/10/2023
								A0908	Incorrect, Inadequate or Imprecise Result or Readings	10/10/2023	10/10/2023
						F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A070101	Failure to Capture	10/10/2023	10/10/2023
								A0705	Battery Problem	10/10/2023	10/10/2023
								A0908	Incorrect, Inadequate or Imprecise Result or Readings	10/10/2023	10/10/2023
						F1905	Device Revision or Replacement	A070101	Failure to Capture	10/10/2023	10/10/2023
								A0705	Battery Problem	10/10/2023	10/10/2023
								A0908	Incorrect, Inadequate or Imprecise Result or Readings	10/10/2023	10/10/2023
14276	2023	All other reportable incidents	36241	Pacemaker/defibrillator lead	CT1684	Cardiac pacemakers and associated devices	E011902	Presyncope	A070909	Oversensing		
									A0712	Pacing Problem		
									A072201	High Impedance		
									A090801	Signal Artifact/Noise		
									A12	Connection Problem		
							E060101	Asystole	A070909	Oversensing		
									A0712	Pacing Problem		
									A072201	High Impedance		
									A090801	Signal Artifact/Noise		
									A12	Connection Problem		
							F08	Hospitalization or Prolonged Hospitalization	A070909	Oversensing		
									A0712	Pacing Problem		
									A072201	High Impedance		
									A090801	Signal Artifact/Noise		
									A12	Connection Problem		
							F1001	Absence of Treatment	A070909	Oversensing		
									A0712	Pacing Problem		
									A072201	High Impedance		
									A090801	Signal Artifact/Noise		
									A12	Connection Problem		
							F1905	Device Revision or Replacement	A070909	Oversensing		
									A0712	Pacing Problem		
									A072201	High Impedance		
									A090801	Signal Artifact/Noise		
									A12	Connection Problem		
14277	2023		47265	Dual-chamber implantable pacemaker, rate-responsive	CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	14/10/2023	14/10/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	14/10/2023	14/10/2023
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	14/10/2023	14/10/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	14/10/2023	14/10/2023
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	14/10/2023	14/10/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	14/10/2023	14/10/2023
14278	2023		47265	Dual-chamber implantable pacemaker, rate-responsive	CT141	Electrodes/Leads and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	14/10/2023	14/10/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	14/10/2023	14/10/2023
					CT145	Cardiovascular devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	14/10/2023	14/10/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	14/10/2023	14/10/2023
					CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	14/10/2023	14/10/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	14/10/2023	14/10/2023
					CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	14/10/2023	14/10/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	14/10/2023	14/10/2023
					CT233	Surgical	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	14/10/2023	14/10/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	14/10/2023	14/10/2023
					CT2406	Implantable	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	14/10/2023	14/10/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	14/10/2023	14/10/2023
					CT2407	Active implantable	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	14/10/2023	14/10/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	14/10/2023	14/10/2023
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	14/10/2023	14/10/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	14/10/2023	14/10/2023
					CT2727	Active implant devices/ancillaries	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	14/10/2023	14/10/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	14/10/2023	14/10/2023
					CT2750	Implantable cardiac electrical stimulators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	14/10/2023	14/10/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	14/10/2023	14/10/2023
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	14/10/2023	14/10/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	14/10/2023	14/10/2023
					CT2985	Active	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	14/10/2023	14/10/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	14/10/2023	14/10/2023
					CT3001	Implantable metal	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	14/10/2023	14/10/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	14/10/2023	14/10/2023
					CT3002	Implantable synthetic polymer	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	14/10/2023	14/10/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	14/10/2023	14/10/2023
					CT3006	Implantable inorganic material	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	14/10/2023	14/10/2023

							F1901	Additional Surgery	A24	No Apparent Device Problem	14/10/2023	14/10/2023
					CT314	Electrically powered	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	14/10/2023	14/10/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	14/10/2023	14/10/2023
					CT321	Long-term surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	14/10/2023	14/10/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	14/10/2023	14/10/2023
					CT449	Generators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	14/10/2023	14/10/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	14/10/2023	14/10/2023
					CT981	Single use	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	14/10/2023	14/10/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	14/10/2023	14/10/2023
					CT983	Surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	14/10/2023	14/10/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	14/10/2023	14/10/2023
14279	2023	All other reportable incidents	35223	Endocardial/interventricular septal pacing lead	CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A0401	Break	27/06/2023	27/06/2023
									A070909	Oversensing	27/06/2023	27/06/2023
							F1901	Additional Surgery	A0401	Break	27/06/2023	27/06/2023
									A070909	Oversensing	27/06/2023	27/06/2023
14280	2023		47265	Dual-chamber implantable pacemaker, rate-responsive	CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	14/10/2023	14/10/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	14/10/2023	14/10/2023
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	14/10/2023	14/10/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	14/10/2023	14/10/2023
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	14/10/2023	14/10/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	14/10/2023	14/10/2023
14281	2023		47265	Dual-chamber implantable pacemaker, rate-responsive	CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	14/10/2023	14/10/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	14/10/2023	14/10/2023
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	14/10/2023	14/10/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	14/10/2023	14/10/2023
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	14/10/2023	14/10/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	14/10/2023	14/10/2023
14282	2023		47265	Dual-chamber implantable pacemaker, rate-responsive	CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	17/10/2023	17/10/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	17/10/2023	17/10/2023
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	17/10/2023	17/10/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	17/10/2023	17/10/2023
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	17/10/2023	17/10/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	17/10/2023	17/10/2023
14283	2023	All other reportable incidents	17882	Physiologic-monitoring defibrillation system	CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070802	Intermittent Loss of Power		
							F27	Problem Identified During Non-Clinical Procedure	A070802	Intermittent Loss of Power		
14284	2023	All other reportable incidents	47265	Dual-chamber implantable pacemaker, rate-responsive	CT141	Electrodes/Leads and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A0908	Incorrect, Inadequate or Imprecise Result or Readings	10/07/2023	10/07/2023
							F08	Hospitalization or Prolonged Hospitalization	A0908	Incorrect, Inadequate or Imprecise Result or Readings	10/07/2023	10/07/2023
							F1905	Device Revision or Replacement	A0908	Incorrect, Inadequate or Imprecise Result or Readings	10/07/2023	10/07/2023
					CT145	Cardiovascular devices	E2403	No Clinical Signs, Symptoms or Conditions	A0908	Incorrect, Inadequate or Imprecise Result or Readings	10/07/2023	10/07/2023
							F08	Hospitalization or Prolonged Hospitalization	A0908	Incorrect, Inadequate or Imprecise Result or Readings	10/07/2023	10/07/2023
							F1905	Device Revision or Replacement	A0908	Incorrect, Inadequate or Imprecise Result or Readings	10/07/2023	10/07/2023
					CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A0908	Incorrect, Inadequate or Imprecise Result or Readings	10/07/2023	10/07/2023
							F08	Hospitalization or Prolonged Hospitalization	A0908	Incorrect, Inadequate or Imprecise Result or Readings	10/07/2023	10/07/2023
							F1905	Device Revision or Replacement	A0908	Incorrect, Inadequate or Imprecise Result or Readings	10/07/2023	10/07/2023
					CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A0908	Incorrect, Inadequate or Imprecise Result or Readings	10/07/2023	10/07/2023
							F08	Hospitalization or Prolonged Hospitalization	A0908	Incorrect, Inadequate or Imprecise Result or Readings	10/07/2023	10/07/2023
							F1905	Device Revision or Replacement	A0908	Incorrect, Inadequate or Imprecise Result or Readings	10/07/2023	10/07/2023
					CT233	Surgical	E2403	No Clinical Signs, Symptoms or Conditions	A0908	Incorrect, Inadequate or Imprecise Result or Readings	10/07/2023	10/07/2023
							F08	Hospitalization or Prolonged Hospitalization	A0908	Incorrect, Inadequate or Imprecise Result or Readings	10/07/2023	10/07/2023
							F1905	Device Revision or Replacement	A0908	Incorrect, Inadequate or Imprecise Result or Readings	10/07/2023	10/07/2023
					CT2406	Implantable	E2403	No Clinical Signs, Symptoms or Conditions	A0908	Incorrect, Inadequate or Imprecise Result or Readings	10/07/2023	10/07/2023
							F08	Hospitalization or Prolonged Hospitalization	A0908	Incorrect, Inadequate or Imprecise Result or Readings	10/07/2023	10/07/2023
							F1905	Device Revision or Replacement	A0908	Incorrect, Inadequate or Imprecise Result or Readings	10/07/2023	10/07/2023
					CT2407	Active implantable	E2403	No Clinical Signs, Symptoms or Conditions	A0908	Incorrect, Inadequate or Imprecise Result or Readings	10/07/2023	10/07/2023
							F08	Hospitalization or Prolonged Hospitalization	A0908	Incorrect, Inadequate or Imprecise Result or Readings	10/07/2023	10/07/2023
							F1905	Device Revision or Replacement	A0908	Incorrect, Inadequate or Imprecise Result or Readings	10/07/2023	10/07/2023
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A0908	Incorrect, Inadequate or Imprecise Result or Readings	10/07/2023	10/07/2023
							F08	Hospitalization or Prolonged Hospitalization	A0908	Incorrect, Inadequate or Imprecise Result or Readings	10/07/2023	10/07/2023
							F1905	Device Revision or Replacement	A0908	Incorrect, Inadequate or Imprecise Result or Readings	10/07/2023	10/07/2023
					CT2727	Active implant devices/ancillaries	E2403	No Clinical Signs, Symptoms or Conditions	A0908	Incorrect, Inadequate or Imprecise Result or Readings	10/07/2023	10/07/2023
							F08	Hospitalization or Prolonged Hospitalization	A0908	Incorrect, Inadequate or Imprecise Result or Readings	10/07/2023	10/07/2023
							F1905	Device Revision or Replacement	A0908	Incorrect, Inadequate or Imprecise Result or Readings	10/07/2023	10/07/2023
					CT2750	Implantable cardiac electrical stimulators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A0908	Incorrect, Inadequate or Imprecise Result or Readings	10/07/2023	10/07/2023
							F08	Hospitalization or Prolonged Hospitalization	A0908	Incorrect, Inadequate or Imprecise Result or Readings	10/07/2023	10/07/2023
							F1905	Device Revision or Replacement	A0908	Incorrect, Inadequate or Imprecise Result or Readings	10/07/2023	10/07/2023

					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A0908	Incorrect, Inadequate or Imprecise Result or Readings	10/07/2023	10/07/2023
							F08	Hospitalization or Prolonged Hospitalization	A0908	Incorrect, Inadequate or Imprecise Result or Readings	10/07/2023	10/07/2023
							F1905	Device Revision or Replacement	A0908	Incorrect, Inadequate or Imprecise Result or Readings	10/07/2023	10/07/2023
					CT2985	Active	E2403	No Clinical Signs, Symptoms or Conditions	A0908	Incorrect, Inadequate or Imprecise Result or Readings	10/07/2023	10/07/2023
							F08	Hospitalization or Prolonged Hospitalization	A0908	Incorrect, Inadequate or Imprecise Result or Readings	10/07/2023	10/07/2023
							F1905	Device Revision or Replacement	A0908	Incorrect, Inadequate or Imprecise Result or Readings	10/07/2023	10/07/2023
					CT3001	Implantable metal	E2403	No Clinical Signs, Symptoms or Conditions	A0908	Incorrect, Inadequate or Imprecise Result or Readings	10/07/2023	10/07/2023
							F08	Hospitalization or Prolonged Hospitalization	A0908	Incorrect, Inadequate or Imprecise Result or Readings	10/07/2023	10/07/2023
							F1905	Device Revision or Replacement	A0908	Incorrect, Inadequate or Imprecise Result or Readings	10/07/2023	10/07/2023
					CT3002	Implantable synthetic polymer	E2403	No Clinical Signs, Symptoms or Conditions	A0908	Incorrect, Inadequate or Imprecise Result or Readings	10/07/2023	10/07/2023
							F08	Hospitalization or Prolonged Hospitalization	A0908	Incorrect, Inadequate or Imprecise Result or Readings	10/07/2023	10/07/2023
							F1905	Device Revision or Replacement	A0908	Incorrect, Inadequate or Imprecise Result or Readings	10/07/2023	10/07/2023
					CT3006	Implantable inorganic material	E2403	No Clinical Signs, Symptoms or Conditions	A0908	Incorrect, Inadequate or Imprecise Result or Readings	10/07/2023	10/07/2023
							F08	Hospitalization or Prolonged Hospitalization	A0908	Incorrect, Inadequate or Imprecise Result or Readings	10/07/2023	10/07/2023
							F1905	Device Revision or Replacement	A0908	Incorrect, Inadequate or Imprecise Result or Readings	10/07/2023	10/07/2023
					CT314	Electrically powered	E2403	No Clinical Signs, Symptoms or Conditions	A0908	Incorrect, Inadequate or Imprecise Result or Readings	10/07/2023	10/07/2023
							F08	Hospitalization or Prolonged Hospitalization	A0908	Incorrect, Inadequate or Imprecise Result or Readings	10/07/2023	10/07/2023
							F1905	Device Revision or Replacement	A0908	Incorrect, Inadequate or Imprecise Result or Readings	10/07/2023	10/07/2023
					CT321	Long-term surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A0908	Incorrect, Inadequate or Imprecise Result or Readings	10/07/2023	10/07/2023
							F08	Hospitalization or Prolonged Hospitalization	A0908	Incorrect, Inadequate or Imprecise Result or Readings	10/07/2023	10/07/2023
							F1905	Device Revision or Replacement	A0908	Incorrect, Inadequate or Imprecise Result or Readings	10/07/2023	10/07/2023
					CT449	Generators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A0908	Incorrect, Inadequate or Imprecise Result or Readings	10/07/2023	10/07/2023
							F08	Hospitalization or Prolonged Hospitalization	A0908	Incorrect, Inadequate or Imprecise Result or Readings	10/07/2023	10/07/2023
							F1905	Device Revision or Replacement	A0908	Incorrect, Inadequate or Imprecise Result or Readings	10/07/2023	10/07/2023
					CT981	Single use	E2403	No Clinical Signs, Symptoms or Conditions	A0908	Incorrect, Inadequate or Imprecise Result or Readings	10/07/2023	10/07/2023
							F08	Hospitalization or Prolonged Hospitalization	A0908	Incorrect, Inadequate or Imprecise Result or Readings	10/07/2023	10/07/2023
							F1905	Device Revision or Replacement	A0908	Incorrect, Inadequate or Imprecise Result or Readings	10/07/2023	10/07/2023
					CT983	Surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A0908	Incorrect, Inadequate or Imprecise Result or Readings	10/07/2023	10/07/2023
							F08	Hospitalization or Prolonged Hospitalization	A0908	Incorrect, Inadequate or Imprecise Result or Readings	10/07/2023	10/07/2023
							F1905	Device Revision or Replacement	A0908	Incorrect, Inadequate or Imprecise Result or Readings	10/07/2023	10/07/2023
14285	2023	All other reportable incidents	47265	Dual-chamber implantable pacemaker, rate-responsive	CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A0908	Incorrect, Inadequate or Imprecise Result or Readings	01/08/2024	01/08/2024
							F08	Hospitalization or Prolonged Hospitalization	A0908	Incorrect, Inadequate or Imprecise Result or Readings	01/08/2024	01/08/2024
							F1905	Device Revision or Replacement	A0908	Incorrect, Inadequate or Imprecise Result or Readings	01/08/2024	01/08/2024
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A0908	Incorrect, Inadequate or Imprecise Result or Readings	01/08/2024	01/08/2024
							F08	Hospitalization or Prolonged Hospitalization	A0908	Incorrect, Inadequate or Imprecise Result or Readings	01/08/2024	01/08/2024
							F1905	Device Revision or Replacement	A0908	Incorrect, Inadequate or Imprecise Result or Readings	01/08/2024	01/08/2024
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A0908	Incorrect, Inadequate or Imprecise Result or Readings	01/08/2024	01/08/2024
							F08	Hospitalization or Prolonged Hospitalization	A0908	Incorrect, Inadequate or Imprecise Result or Readings	01/08/2024	01/08/2024
							F1905	Device Revision or Replacement	A0908	Incorrect, Inadequate or Imprecise Result or Readings	01/08/2024	01/08/2024
14286	2023	All other reportable incidents	47265	Dual-chamber implantable pacemaker, rate-responsive	CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A05	Mechanical Problem	28/09/2023	28/09/2023
							F08	Hospitalization or Prolonged Hospitalization	A05	Mechanical Problem	28/09/2023	28/09/2023
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A05	Mechanical Problem	28/09/2023	28/09/2023
							F08	Hospitalization or Prolonged Hospitalization	A05	Mechanical Problem	28/09/2023	28/09/2023
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A05	Mechanical Problem	28/09/2023	28/09/2023
							F08	Hospitalization or Prolonged Hospitalization	A05	Mechanical Problem	28/09/2023	28/09/2023
14287	2023	All other reportable incidents	47263	Cardiac resynchronization therapy implantable pacemaker	CT1685	Cardiac pacemakers	E011902	Presyncope	A070909	Oversensing	12/10/2023	12/10/2023
									A0712	Pacing Problem	12/10/2023	12/10/2023
									A072201	High Impedance	12/10/2023	12/10/2023
									A090801	Signal Artifact/Noise	12/10/2023	12/10/2023
									A12	Connection Problem	12/10/2023	12/10/2023
							E060101	Asystole	A070909	Oversensing	12/10/2023	12/10/2023
									A0712	Pacing Problem	12/10/2023	12/10/2023
									A072201	High Impedance	12/10/2023	12/10/2023
									A090801	Signal Artifact/Noise	12/10/2023	12/10/2023
									A12	Connection Problem	12/10/2023	12/10/2023
							F08	Hospitalization or Prolonged Hospitalization	A070909	Oversensing	12/10/2023	12/10/2023
									A0712	Pacing Problem	12/10/2023	12/10/2023
									A072201	High Impedance	12/10/2023	12/10/2023

									A090801	Signal Artifact/Noise	12/10/2023	12/10/2023
									A12	Connection Problem	12/10/2023	12/10/2023
							F1001	Absence of Treatment	A070909	Oversensing	12/10/2023	12/10/2023
									A0712	Pacing Problem	12/10/2023	12/10/2023
									A072201	High Impedance	12/10/2023	12/10/2023
									A090801	Signal Artifact/Noise	12/10/2023	12/10/2023
									A12	Connection Problem	12/10/2023	12/10/2023
							F1905	Device Revision or Replacement	A070909	Oversensing	12/10/2023	12/10/2023
									A0712	Pacing Problem	12/10/2023	12/10/2023
									A072201	High Impedance	12/10/2023	12/10/2023
									A090801	Signal Artifact/Noise	12/10/2023	12/10/2023
									A12	Connection Problem	12/10/2023	12/10/2023
					CT257	Implantable cardiac pacemakers	E011902	Presyncope	A070909	Oversensing	12/10/2023	12/10/2023
									A0712	Pacing Problem	12/10/2023	12/10/2023
									A072201	High Impedance	12/10/2023	12/10/2023
									A090801	Signal Artifact/Noise	12/10/2023	12/10/2023
									A12	Connection Problem	12/10/2023	12/10/2023
							E060101	Asystole	A070909	Oversensing	12/10/2023	12/10/2023
									A0712	Pacing Problem	12/10/2023	12/10/2023
									A072201	High Impedance	12/10/2023	12/10/2023
									A090801	Signal Artifact/Noise	12/10/2023	12/10/2023
									A12	Connection Problem	12/10/2023	12/10/2023
							F08	Hospitalization or Prolonged Hospitalization	A070909	Oversensing	12/10/2023	12/10/2023
									A0712	Pacing Problem	12/10/2023	12/10/2023
									A072201	High Impedance	12/10/2023	12/10/2023
									A090801	Signal Artifact/Noise	12/10/2023	12/10/2023
									A12	Connection Problem	12/10/2023	12/10/2023
							F1001	Absence of Treatment	A070909	Oversensing	12/10/2023	12/10/2023
									A0712	Pacing Problem	12/10/2023	12/10/2023
									A072201	High Impedance	12/10/2023	12/10/2023
									A090801	Signal Artifact/Noise	12/10/2023	12/10/2023
									A12	Connection Problem	12/10/2023	12/10/2023
							F1905	Device Revision or Replacement	A070909	Oversensing	12/10/2023	12/10/2023
									A0712	Pacing Problem	12/10/2023	12/10/2023
									A072201	High Impedance	12/10/2023	12/10/2023
									A090801	Signal Artifact/Noise	12/10/2023	12/10/2023
									A12	Connection Problem	12/10/2023	12/10/2023
					CT2866	Implantable cardiac electrical stimulators	E011902	Presyncope	A070909	Oversensing	12/10/2023	12/10/2023
									A0712	Pacing Problem	12/10/2023	12/10/2023
									A072201	High Impedance	12/10/2023	12/10/2023
									A090801	Signal Artifact/Noise	12/10/2023	12/10/2023
									A12	Connection Problem	12/10/2023	12/10/2023
							E060101	Asystole	A070909	Oversensing	12/10/2023	12/10/2023
									A0712	Pacing Problem	12/10/2023	12/10/2023
									A072201	High Impedance	12/10/2023	12/10/2023
									A090801	Signal Artifact/Noise	12/10/2023	12/10/2023
									A12	Connection Problem	12/10/2023	12/10/2023
							F08	Hospitalization or Prolonged Hospitalization	A070909	Oversensing	12/10/2023	12/10/2023
									A0712	Pacing Problem	12/10/2023	12/10/2023
									A072201	High Impedance	12/10/2023	12/10/2023
									A090801	Signal Artifact/Noise	12/10/2023	12/10/2023
									A12	Connection Problem	12/10/2023	12/10/2023
							F1001	Absence of Treatment	A070909	Oversensing	12/10/2023	12/10/2023
									A0712	Pacing Problem	12/10/2023	12/10/2023
									A072201	High Impedance	12/10/2023	12/10/2023
									A090801	Signal Artifact/Noise	12/10/2023	12/10/2023
									A12	Connection Problem	12/10/2023	12/10/2023
							F1905	Device Revision or Replacement	A070909	Oversensing	12/10/2023	12/10/2023
									A0712	Pacing Problem	12/10/2023	12/10/2023
									A072201	High Impedance	12/10/2023	12/10/2023
									A090801	Signal Artifact/Noise	12/10/2023	12/10/2023
									A12	Connection Problem	12/10/2023	12/10/2023
14288	2023		47265	Dual-chamber implantable pacemaker, rate-responsive	CT141	Electrodes/Leads and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A0403	Explosion		
									A0501	Detachment of Device or Device Component		
									A1411	Pressure Problem		
							F26	No Health Consequences or Impact	A0403	Explosion		
									A0501	Detachment of Device or Device Component		
									A1411	Pressure Problem		
					CT145	Cardiovascular devices	E2403	No Clinical Signs, Symptoms or Conditions	A0403	Explosion		
									A0501	Detachment of Device or Device Component		
									A1411	Pressure Problem		
							F26	No Health Consequences or Impact	A0403	Explosion		
									A0501	Detachment of Device or Device Component		
									A1411	Pressure Problem		
					CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A0403	Explosion		
									A0501	Detachment of Device or Device Component		
									A1411	Pressure Problem		
							F26	No Health Consequences or Impact	A0403	Explosion		
									A0501	Detachment of Device or Device Component		
									A1411	Pressure Problem		
					CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A0403	Explosion		
									A0501	Detachment of Device or Device Component		
									A1411	Pressure Problem		
							F26	No Health Consequences or Impact	A0403	Explosion		
									A0501	Detachment of Device or Device Component		
									A1411	Pressure Problem		
					CT233	Surgical	E2403	No Clinical Signs, Symptoms or Conditions	A0403	Explosion		
									A0501	Detachment of Device or Device Component		

								A1411	Pressure Problem		
						F26	No Health Consequences or Impact	A0403	Explosion		
								A0501	Detachment of Device or Device Component		
								A1411	Pressure Problem		
					CT2406	Implantable	E2403	No Clinical Signs, Symptoms or Conditions	A0403	Explosion		
									A0501	Detachment of Device or Device Component		
									A1411	Pressure Problem		
							F26	No Health Consequences or Impact	A0403	Explosion		
									A0501	Detachment of Device or Device Component		
									A1411	Pressure Problem		
					CT2407	Active implantable	E2403	No Clinical Signs, Symptoms or Conditions	A0403	Explosion		
									A0501	Detachment of Device or Device Component		
									A1411	Pressure Problem		
							F26	No Health Consequences or Impact	A0403	Explosion		
									A0501	Detachment of Device or Device Component		
									A1411	Pressure Problem		
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A0403	Explosion		
									A0501	Detachment of Device or Device Component		
									A1411	Pressure Problem		
							F26	No Health Consequences or Impact	A0403	Explosion		
									A0501	Detachment of Device or Device Component		
									A1411	Pressure Problem		
					CT2727	Active implant devices/ancillaries	E2403	No Clinical Signs, Symptoms or Conditions	A0403	Explosion		
									A0501	Detachment of Device or Device Component		
									A1411	Pressure Problem		
							F26	No Health Consequences or Impact	A0403	Explosion		
									A0501	Detachment of Device or Device Component		
									A1411	Pressure Problem		
					CT2750	Implantable cardiac electrical stimulators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A0403	Explosion		
									A0501	Detachment of Device or Device Component		
									A1411	Pressure Problem		
							F26	No Health Consequences or Impact	A0403	Explosion		
									A0501	Detachment of Device or Device Component		
									A1411	Pressure Problem		
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A0403	Explosion		
									A0501	Detachment of Device or Device Component		
									A1411	Pressure Problem		
							F26	No Health Consequences or Impact	A0403	Explosion		
									A0501	Detachment of Device or Device Component		
									A1411	Pressure Problem		
					CT2985	Active	E2403	No Clinical Signs, Symptoms or Conditions	A0403	Explosion		
									A0501	Detachment of Device or Device Component		
									A1411	Pressure Problem		
							F26	No Health Consequences or Impact	A0403	Explosion		
									A0501	Detachment of Device or Device Component		
									A1411	Pressure Problem		
					CT3001	Implantable metal	E2403	No Clinical Signs, Symptoms or Conditions	A0403	Explosion		
									A0501	Detachment of Device or Device Component		
									A1411	Pressure Problem		
							F26	No Health Consequences or Impact	A0403	Explosion		
									A0501	Detachment of Device or Device Component		
									A1411	Pressure Problem		
					CT3002	Implantable synthetic polymer	E2403	No Clinical Signs, Symptoms or Conditions	A0403	Explosion		
									A0501	Detachment of Device or Device Component		
									A1411	Pressure Problem		
							F26	No Health Consequences or Impact	A0403	Explosion		
									A0501	Detachment of Device or Device Component		
									A1411	Pressure Problem		
					CT3006	Implantable inorganic material	E2403	No Clinical Signs, Symptoms or Conditions	A0403	Explosion		
									A0501	Detachment of Device or Device Component		
									A1411	Pressure Problem		
							F26	No Health Consequences or Impact	A0403	Explosion		
									A0501	Detachment of Device or Device Component		
									A1411	Pressure Problem		
					CT314	Electrically powered	E2403	No Clinical Signs, Symptoms or Conditions	A0403	Explosion		
									A0501	Detachment of Device or Device Component		
									A1411	Pressure Problem		
							F26	No Health Consequences or Impact	A0403	Explosion		
									A0501	Detachment of Device or Device Component		
									A1411	Pressure Problem		
					CT321	Long-term surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A0403	Explosion		
									A0501	Detachment of Device or Device Component		
									A1411	Pressure Problem		
							F26	No Health Consequences or Impact	A0403	Explosion		
									A0501	Detachment of Device or Device Component		
									A1411	Pressure Problem		
					CT449	Generators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A0403	Explosion		
									A0501	Detachment of Device or Device Component		
									A1411	Pressure Problem		
							F26	No Health Consequences or Impact	A0403	Explosion		
									A0501	Detachment of Device or Device Component		
									A1411	Pressure Problem		
					CT981	Single use	E2403	No Clinical Signs, Symptoms or Conditions	A0403	Explosion		
									A0501	Detachment of Device or Device Component		
									A1411	Pressure Problem		
							F26	No Health Consequences or Impact	A0403	Explosion		
									A0501	Detachment of Device or Device Component		
									A1411	Pressure Problem		

					CT983	Surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A0403	Explosion			
									A0501	Detachment of Device or Device Component			
									A1411	Pressure Problem			
							F26	No Health Consequences or Impact	A0403	Explosion			
									A0501	Detachment of Device or Device Component			
									A1411	Pressure Problem			
14289	2023		47265	Dual-chamber implantable pacemaker, rate-responsive	CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	17/10/2023	17/10/2023	
							F1901	Additional Surgery	A24	No Apparent Device Problem	17/10/2023	17/10/2023	
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	17/10/2023	17/10/2023	
							F1901	Additional Surgery	A24	No Apparent Device Problem	17/10/2023	17/10/2023	
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	17/10/2023	17/10/2023	
							F1901	Additional Surgery	A24	No Apparent Device Problem	17/10/2023	17/10/2023	
14290	2023		47265	Dual-chamber implantable pacemaker, rate-responsive	CT141	Electrodes/Leads and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	19/10/2023	19/10/2023	
							F1901	Additional Surgery	A24	No Apparent Device Problem	19/10/2023	19/10/2023	
					CT145	Cardiovascular devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	19/10/2023	19/10/2023	
							F1901	Additional Surgery	A24	No Apparent Device Problem	19/10/2023	19/10/2023	
					CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	19/10/2023	19/10/2023	
							F1901	Additional Surgery	A24	No Apparent Device Problem	19/10/2023	19/10/2023	
					CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	19/10/2023	19/10/2023	
							F1901	Additional Surgery	A24	No Apparent Device Problem	19/10/2023	19/10/2023	
					CT233	Surgical	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	19/10/2023	19/10/2023	
							F1901	Additional Surgery	A24	No Apparent Device Problem	19/10/2023	19/10/2023	
					CT2406	Implantable	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	19/10/2023	19/10/2023	
							F1901	Additional Surgery	A24	No Apparent Device Problem	19/10/2023	19/10/2023	
					CT2407	Active implantable	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	19/10/2023	19/10/2023	
							F1901	Additional Surgery	A24	No Apparent Device Problem	19/10/2023	19/10/2023	
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	19/10/2023	19/10/2023	
							F1901	Additional Surgery	A24	No Apparent Device Problem	19/10/2023	19/10/2023	
					CT2727	Active implant devices/ancillaries	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	19/10/2023	19/10/2023	
							F1901	Additional Surgery	A24	No Apparent Device Problem	19/10/2023	19/10/2023	
					CT2750	Implantable cardiac electrical stimulators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	19/10/2023	19/10/2023	
							F1901	Additional Surgery	A24	No Apparent Device Problem	19/10/2023	19/10/2023	
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	19/10/2023	19/10/2023	
							F1901	Additional Surgery	A24	No Apparent Device Problem	19/10/2023	19/10/2023	
					CT2985	Active	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	19/10/2023	19/10/2023	
							F1901	Additional Surgery	A24	No Apparent Device Problem	19/10/2023	19/10/2023	
					CT3001	Implantable metal	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	19/10/2023	19/10/2023	
							F1901	Additional Surgery	A24	No Apparent Device Problem	19/10/2023	19/10/2023	
					CT3002	Implantable synthetic polymer	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	19/10/2023	19/10/2023	
							F1901	Additional Surgery	A24	No Apparent Device Problem	19/10/2023	19/10/2023	
					CT3006	Implantable inorganic material	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	19/10/2023	19/10/2023	
							F1901	Additional Surgery	A24	No Apparent Device Problem	19/10/2023	19/10/2023	
					CT314	Electrically powered	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	19/10/2023	19/10/2023	
							F1901	Additional Surgery	A24	No Apparent Device Problem	19/10/2023	19/10/2023	
					CT321	Long-term surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	19/10/2023	19/10/2023	
							F1901	Additional Surgery	A24	No Apparent Device Problem	19/10/2023	19/10/2023	
					CT449	Generators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	19/10/2023	19/10/2023	
							F1901	Additional Surgery	A24	No Apparent Device Problem	19/10/2023	19/10/2023	
					CT981	Single use	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	19/10/2023	19/10/2023	
							F1901	Additional Surgery	A24	No Apparent Device Problem	19/10/2023	19/10/2023	
					CT983	Surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	19/10/2023	19/10/2023	
							F1901	Additional Surgery	A24	No Apparent Device Problem	19/10/2023	19/10/2023	
14291	2023		47265	Dual-chamber implantable pacemaker, rate-responsive	CT141	Electrodes/Leads and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	19/10/2023	19/10/2023	
							F1901	Additional Surgery	A24	No Apparent Device Problem	19/10/2023	19/10/2023	
					CT145	Cardiovascular devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	19/10/2023	19/10/2023	
							F1901	Additional Surgery	A24	No Apparent Device Problem	19/10/2023	19/10/2023	
					CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	19/10/2023	19/10/2023	
							F1901	Additional Surgery	A24	No Apparent Device Problem	19/10/2023	19/10/2023	
					CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	19/10/2023	19/10/2023	
							F1901	Additional Surgery	A24	No Apparent Device Problem	19/10/2023	19/10/2023	
					CT233	Surgical	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	19/10/2023	19/10/2023	
							F1901	Additional Surgery	A24	No Apparent Device Problem	19/10/2023	19/10/2023	
					CT2406	Implantable	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	19/10/2023	19/10/2023	
							F1901	Additional Surgery	A24	No Apparent Device Problem	19/10/2023	19/10/2023	
					CT2407	Active implantable	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	19/10/2023	19/10/2023	
							F1901	Additional Surgery	A24	No Apparent Device Problem	19/10/2023	19/10/2023	
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	19/10/2023	19/10/2023	
							F1901	Additional Surgery	A24	No Apparent Device Problem	19/10/2023	19/10/2023	
					CT2727	Active implant devices/ancillaries	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	19/10/2023	19/10/2023	
							F1901	Additional Surgery	A24	No Apparent Device Problem	19/10/2023	19/10/2023	
					CT2750	Implantable cardiac electrical stimulators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	19/10/2023	19/10/2023	
							F1901	Additional Surgery	A24	No Apparent Device Problem	19/10/2023	19/10/2023	
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	19/10/2023	19/10/2023	
							F1901	Additional Surgery	A24	No Apparent Device Problem	19/10/2023	19/10/2023	
					CT2985	Active	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	19/10/2023	19/10/2023	
							F1901	Additional Surgery	A24	No Apparent Device Problem	19/10/2023	19/10/2023	
					CT3001	Implantable metal	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	19/10/2023	19/10/2023	
							F1901	Additional Surgery	A24	No Apparent Device Problem	19/10/2023	19/10/2023	
					CT3002	Implantable synthetic polymer	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	19/10/2023	19/10/2023	
							F1901	Additional Surgery	A24	No Apparent Device Problem	19/10/2023	19/10/2023	
					CT3006	Implantable inorganic material	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	19/10/2023	19/10/2023	
							F1901	Additional Surgery	A24	No Apparent Device Problem	19/10/2023	19/10/2023	
					CT314	Electrically powered	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	19/10/2023	19/10/2023	
							F1901	Additional Surgery	A24	No Apparent Device Problem	19/10/2023	19/10/2023	
					CT321	Long-term surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	19/10/2023	19/10/2023	
							F1901	Additional Surgery	A24	No Apparent Device Problem	19/10/2023	19/10/2023	
					CT449	Generators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	19/10/2023	19/10/2023	
							F1901	Additional Surgery	A24	No Apparent Device Problem	19/10/2023	19/10/2023	

					CT981	Single use	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	19/10/2023	19/10/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	19/10/2023	19/10/2023
					CT983	Surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	19/10/2023	19/10/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	19/10/2023	19/10/2023
14292	2023		47265	Dual-chamber implantable pacemaker, rate-responsive	CT141	Electrodes/Leads and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	12/10/2023	12/10/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	12/10/2023	12/10/2023
					CT145	Cardiovascular devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	12/10/2023	12/10/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	12/10/2023	12/10/2023
					CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	12/10/2023	12/10/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	12/10/2023	12/10/2023
					CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	12/10/2023	12/10/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	12/10/2023	12/10/2023
					CT233	Surgical	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	12/10/2023	12/10/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	12/10/2023	12/10/2023
					CT2406	Implantable	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	12/10/2023	12/10/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	12/10/2023	12/10/2023
					CT2407	Active implantable	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	12/10/2023	12/10/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	12/10/2023	12/10/2023
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	12/10/2023	12/10/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	12/10/2023	12/10/2023
					CT2727	Active implant devices/ancillaries	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	12/10/2023	12/10/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	12/10/2023	12/10/2023
					CT2750	Implantable cardiac electrical stimulators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	12/10/2023	12/10/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	12/10/2023	12/10/2023
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	12/10/2023	12/10/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	12/10/2023	12/10/2023
					CT2985	Active	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	12/10/2023	12/10/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	12/10/2023	12/10/2023
					CT3001	Implantable metal	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	12/10/2023	12/10/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	12/10/2023	12/10/2023
					CT3002	Implantable synthetic polymer	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	12/10/2023	12/10/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	12/10/2023	12/10/2023
					CT3006	Implantable inorganic material	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	12/10/2023	12/10/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	12/10/2023	12/10/2023
					CT314	Electrically powered	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	12/10/2023	12/10/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	12/10/2023	12/10/2023
					CT321	Long-term surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	12/10/2023	12/10/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	12/10/2023	12/10/2023
					CT449	Generators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	12/10/2023	12/10/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	12/10/2023	12/10/2023
					CT981	Single use	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	12/10/2023	12/10/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	12/10/2023	12/10/2023
					CT983	Surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	12/10/2023	12/10/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	12/10/2023	12/10/2023
14293	2023		47265	Dual-chamber implantable pacemaker, rate-responsive	CT141	Electrodes/Leads and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem		
							F1901	Additional Surgery	A24	No Apparent Device Problem		
					CT145	Cardiovascular devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem		
							F1901	Additional Surgery	A24	No Apparent Device Problem		
					CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem		
							F1901	Additional Surgery	A24	No Apparent Device Problem		
					CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem		
							F1901	Additional Surgery	A24	No Apparent Device Problem		
					CT233	Surgical	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem		
							F1901	Additional Surgery	A24	No Apparent Device Problem		
					CT2406	Implantable	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem		
							F1901	Additional Surgery	A24	No Apparent Device Problem		
					CT2407	Active implantable	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem		
							F1901	Additional Surgery	A24	No Apparent Device Problem		
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem		
							F1901	Additional Surgery	A24	No Apparent Device Problem		
					CT2727	Active implant devices/ancillaries	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem		
							F1901	Additional Surgery	A24	No Apparent Device Problem		
					CT2750	Implantable cardiac electrical stimulators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem		
							F1901	Additional Surgery	A24	No Apparent Device Problem		
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem		
							F1901	Additional Surgery	A24	No Apparent Device Problem		
					CT2985	Active	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem		
							F1901	Additional Surgery	A24	No Apparent Device Problem		
					CT3001	Implantable metal	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem		
							F1901	Additional Surgery	A24	No Apparent Device Problem		
					CT3002	Implantable synthetic polymer	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem		
							F1901	Additional Surgery	A24	No Apparent Device Problem		
					CT3006	Implantable inorganic material	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem		
							F1901	Additional Surgery	A24	No Apparent Device Problem		
					CT314	Electrically powered	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem		
							F1901	Additional Surgery	A24	No Apparent Device Problem		
					CT321	Long-term surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem		
							F1901	Additional Surgery	A24	No Apparent Device Problem		
					CT449	Generators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem		
							F1901	Additional Surgery	A24	No Apparent Device Problem		
					CT981	Single use	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem		
							F1901	Additional Surgery	A24	No Apparent Device Problem		
					CT983	Surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem		
							F1901	Additional Surgery	A24	No Apparent Device Problem		
14294	2023		47265	Dual-chamber implantable pacemaker, rate-responsive	CT141	Electrodes/Leads and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem		
							F1901	Additional Surgery	A24	No Apparent Device Problem		
					CT145	Cardiovascular devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem		
							F1901	Additional Surgery	A24	No Apparent Device Problem		

					CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem		
							F1901	Additional Surgery	A24	No Apparent Device Problem		
					CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem		
							F1901	Additional Surgery	A24	No Apparent Device Problem		
					CT233	Surgical	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem		
							F1901	Additional Surgery	A24	No Apparent Device Problem		
					CT2406	Implantable	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem		
							F1901	Additional Surgery	A24	No Apparent Device Problem		
					CT2407	Active implantable	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem		
							F1901	Additional Surgery	A24	No Apparent Device Problem		
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem		
							F1901	Additional Surgery	A24	No Apparent Device Problem		
					CT2727	Active implant devices/ancillaries	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem		
							F1901	Additional Surgery	A24	No Apparent Device Problem		
					CT2750	Implantable cardiac electrical stimulators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem		
							F1901	Additional Surgery	A24	No Apparent Device Problem		
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem		
							F1901	Additional Surgery	A24	No Apparent Device Problem		
					CT2985	Active	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem		
							F1901	Additional Surgery	A24	No Apparent Device Problem		
					CT3001	Implantable metal	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem		
							F1901	Additional Surgery	A24	No Apparent Device Problem		
					CT3002	Implantable synthetic polymer	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem		
							F1901	Additional Surgery	A24	No Apparent Device Problem		
					CT3006	Implantable inorganic material	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem		
							F1901	Additional Surgery	A24	No Apparent Device Problem		
					CT314	Electrically powered	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem		
							F1901	Additional Surgery	A24	No Apparent Device Problem		
					CT321	Long-term surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem		
							F1901	Additional Surgery	A24	No Apparent Device Problem		
					CT449	Generators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem		
							F1901	Additional Surgery	A24	No Apparent Device Problem		
					CT981	Single use	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem		
							F1901	Additional Surgery	A24	No Apparent Device Problem		
					CT983	Surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem		
							F1901	Additional Surgery	A24	No Apparent Device Problem		
14295	2023	All other reportable incidents	35852	Single-chamber implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E2104	Electric Shock	A0709	Device Sensing Problem		
									A070909	Oversensing		
									A071302	Inappropriate/Inadequate Shock/Stimulation		
									A1301	Failure to Read Input Signal		
									A23	Use of Device Problem		
							F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A0709	Device Sensing Problem		
									A070909	Oversensing		
									A071302	Inappropriate/Inadequate Shock/Stimulation		
									A1301	Failure to Read Input Signal		
									A23	Use of Device Problem		
							F25	Unanticipated Adverse Device Effect	A0709	Device Sensing Problem		
									A070909	Oversensing		
									A071302	Inappropriate/Inadequate Shock/Stimulation		
									A1301	Failure to Read Input Signal		
									A23	Use of Device Problem		
14296	2023	All other reportable incidents	35852	Single-chamber implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E2104	Electric Shock	A0709	Device Sensing Problem		
									A070909	Oversensing		
									A071302	Inappropriate/Inadequate Shock/Stimulation		
							F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A0709	Device Sensing Problem		
									A070909	Oversensing		
									A071302	Inappropriate/Inadequate Shock/Stimulation		
14297	2023	All other reportable incidents	37265	Dual-chamber implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E0402	Hypersensitivity/Allergic Reaction	A24	No Apparent Device Problem	01/04/2023	01/04/2023
							F1903	Device Explantation	A24	No Apparent Device Problem	01/04/2023	01/04/2023
14298	2023		35223	Endocardial/interventricular septal pacing lead	CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A0401	Break	12/10/2023	12/10/2023
									A051201	Device Dislodged or Dislocated	12/10/2023	12/10/2023
									A070101	Failure to Capture	12/10/2023	12/10/2023
									A072201	High Impedance	12/10/2023	12/10/2023
							F1901	Additional Surgery	A0401	Break	12/10/2023	12/10/2023
									A051201	Device Dislodged or Dislocated	12/10/2023	12/10/2023
									A070101	Failure to Capture	12/10/2023	12/10/2023
									A072201	High Impedance	12/10/2023	12/10/2023
14299	2023		35223	Endocardial/interventricular septal pacing lead	CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A0401	Break	12/10/2023	12/10/2023
							F1908	Prolonged Surgery	A0401	Break	12/10/2023	12/10/2023
14300	2023		47265	Dual-chamber implantable pacemaker, rate-responsive	CT141	Electrodes/Leads and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	12/10/2023	12/10/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	12/10/2023	12/10/2023
					CT145	Cardiovascular devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	12/10/2023	12/10/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	12/10/2023	12/10/2023
					CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	12/10/2023	12/10/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	12/10/2023	12/10/2023
					CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	12/10/2023	12/10/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	12/10/2023	12/10/2023
					CT233	Surgical	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	12/10/2023	12/10/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	12/10/2023	12/10/2023
					CT2406	Implantable	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	12/10/2023	12/10/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	12/10/2023	12/10/2023
					CT2407	Active implantable	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	12/10/2023	12/10/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	12/10/2023	12/10/2023
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	12/10/2023	12/10/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	12/10/2023	12/10/2023
					CT2727	Active implant devices/ancillaries	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	12/10/2023	12/10/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	12/10/2023	12/10/2023
					CT2750	Implantable cardiac electrical stimulators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	12/10/2023	12/10/2023

							F1901	Additional Surgery	A24	No Apparent Device Problem	12/10/2023	12/10/2023
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	12/10/2023	12/10/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	12/10/2023	12/10/2023
					CT2985	Active	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	12/10/2023	12/10/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	12/10/2023	12/10/2023
					CT3001	Implantable metal	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	12/10/2023	12/10/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	12/10/2023	12/10/2023
					CT3002	Implantable synthetic polymer	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	12/10/2023	12/10/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	12/10/2023	12/10/2023
					CT3006	Implantable inorganic material	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	12/10/2023	12/10/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	12/10/2023	12/10/2023
					CT314	Electrically powered	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	12/10/2023	12/10/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	12/10/2023	12/10/2023
					CT321	Long-term surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	12/10/2023	12/10/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	12/10/2023	12/10/2023
					CT449	Generators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	12/10/2023	12/10/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	12/10/2023	12/10/2023
					CT981	Single use	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	12/10/2023	12/10/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	12/10/2023	12/10/2023
					CT983	Surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	12/10/2023	12/10/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	12/10/2023	12/10/2023
14301	2023	All other reportable incidents	17882	Physiologic-monitoring defibrillation system	CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A0719	Unexpected Shutdown		
							F27	Problem Identified During Non-Clinical Procedure	A0719	Unexpected Shutdown		
14302	2023		47265	Dual-chamber implantable pacemaker, rate-responsive	CT141	Electrodes/Leads and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	19/10/2023	19/10/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	19/10/2023	19/10/2023
					CT145	Cardiovascular devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	19/10/2023	19/10/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	19/10/2023	19/10/2023
					CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	19/10/2023	19/10/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	19/10/2023	19/10/2023
					CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	19/10/2023	19/10/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	19/10/2023	19/10/2023
					CT233	Surgical	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	19/10/2023	19/10/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	19/10/2023	19/10/2023
					CT2406	Implantable	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	19/10/2023	19/10/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	19/10/2023	19/10/2023
					CT2407	Active implantable	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	19/10/2023	19/10/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	19/10/2023	19/10/2023
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	19/10/2023	19/10/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	19/10/2023	19/10/2023
					CT2727	Active implant devices/ancillaries	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	19/10/2023	19/10/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	19/10/2023	19/10/2023
					CT2750	Implantable cardiac electrical stimulators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	19/10/2023	19/10/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	19/10/2023	19/10/2023
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	19/10/2023	19/10/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	19/10/2023	19/10/2023
					CT2985	Active	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	19/10/2023	19/10/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	19/10/2023	19/10/2023
					CT3001	Implantable metal	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	19/10/2023	19/10/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	19/10/2023	19/10/2023
					CT3002	Implantable synthetic polymer	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	19/10/2023	19/10/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	19/10/2023	19/10/2023
					CT3006	Implantable inorganic material	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	19/10/2023	19/10/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	19/10/2023	19/10/2023
					CT314	Electrically powered	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	19/10/2023	19/10/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	19/10/2023	19/10/2023
					CT321	Long-term surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	19/10/2023	19/10/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	19/10/2023	19/10/2023
					CT449	Generators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	19/10/2023	19/10/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	19/10/2023	19/10/2023
					CT981	Single use	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	19/10/2023	19/10/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	19/10/2023	19/10/2023
					CT983	Surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	19/10/2023	19/10/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	19/10/2023	19/10/2023
14303	2023		47265	Dual-chamber implantable pacemaker, rate-responsive	CT141	Electrodes/Leads and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	19/10/2023	19/10/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	19/10/2023	19/10/2023
					CT145	Cardiovascular devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	19/10/2023	19/10/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	19/10/2023	19/10/2023
					CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	19/10/2023	19/10/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	19/10/2023	19/10/2023
					CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	19/10/2023	19/10/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	19/10/2023	19/10/2023
					CT233	Surgical	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	19/10/2023	19/10/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	19/10/2023	19/10/2023
					CT2406	Implantable	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	19/10/2023	19/10/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	19/10/2023	19/10/2023
					CT2407	Active implantable	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	19/10/2023	19/10/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	19/10/2023	19/10/2023
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	19/10/2023	19/10/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	19/10/2023	19/10/2023
					CT2727	Active implant devices/ancillaries	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	19/10/2023	19/10/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	19/10/2023	19/10/2023
					CT2750	Implantable cardiac electrical stimulators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	19/10/2023	19/10/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	19/10/2023	19/10/2023
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	19/10/2023	19/10/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	19/10/2023	19/10/2023
					CT2985	Active	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	19/10/2023	19/10/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	19/10/2023	19/10/2023
					CT3001	Implantable metal	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	19/10/2023	19/10/2023

							F1901	Additional Surgery	A24	No Apparent Device Problem	19/10/2023	19/10/2023
					CT3002	Implantable synthetic polymer	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	19/10/2023	19/10/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	19/10/2023	19/10/2023
					CT3006	Implantable inorganic material	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	19/10/2023	19/10/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	19/10/2023	19/10/2023
					CT314	Electrically powered	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	19/10/2023	19/10/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	19/10/2023	19/10/2023
					CT321	Long-term surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	19/10/2023	19/10/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	19/10/2023	19/10/2023
					CT449	Generators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	19/10/2023	19/10/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	19/10/2023	19/10/2023
					CT981	Single use	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	19/10/2023	19/10/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	19/10/2023	19/10/2023
					CT983	Surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	19/10/2023	19/10/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	19/10/2023	19/10/2023
14304	2023	All other reportable incidents	47265	Dual-chamber implantable pacemaker, rate-responsive	CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem	01/08/2024	01/08/2024
									A070504	Premature Discharge of Battery	01/08/2024	01/08/2024
							F08	Hospitalization or Prolonged Hospitalization	A0705	Battery Problem	01/08/2024	01/08/2024
									A070504	Premature Discharge of Battery	01/08/2024	01/08/2024
							F1905	Device Revision or Replacement	A0705	Battery Problem	01/08/2024	01/08/2024
									A070504	Premature Discharge of Battery	01/08/2024	01/08/2024
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem	01/08/2024	01/08/2024
									A070504	Premature Discharge of Battery	01/08/2024	01/08/2024
							F08	Hospitalization or Prolonged Hospitalization	A0705	Battery Problem	01/08/2024	01/08/2024
									A070504	Premature Discharge of Battery	01/08/2024	01/08/2024
							F1905	Device Revision or Replacement	A0705	Battery Problem	01/08/2024	01/08/2024
									A070504	Premature Discharge of Battery	01/08/2024	01/08/2024
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem	01/08/2024	01/08/2024
									A070504	Premature Discharge of Battery	01/08/2024	01/08/2024
							F08	Hospitalization or Prolonged Hospitalization	A0705	Battery Problem	01/08/2024	01/08/2024
									A070504	Premature Discharge of Battery	01/08/2024	01/08/2024
							F1905	Device Revision or Replacement	A0705	Battery Problem	01/08/2024	01/08/2024
									A070504	Premature Discharge of Battery	01/08/2024	01/08/2024
14305	2023		47265	Dual-chamber implantable pacemaker, rate-responsive	CT141	Electrodes/Leads and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	17/10/2023	17/10/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	17/10/2023	17/10/2023
					CT145	Cardiovascular devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	17/10/2023	17/10/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	17/10/2023	17/10/2023
					CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	17/10/2023	17/10/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	17/10/2023	17/10/2023
					CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	17/10/2023	17/10/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	17/10/2023	17/10/2023
					CT233	Surgical	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	17/10/2023	17/10/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	17/10/2023	17/10/2023
					CT2406	Implantable	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	17/10/2023	17/10/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	17/10/2023	17/10/2023
					CT2407	Active implantable	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	17/10/2023	17/10/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	17/10/2023	17/10/2023
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	17/10/2023	17/10/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	17/10/2023	17/10/2023
					CT2727	Active implant devices/ancillaries	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	17/10/2023	17/10/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	17/10/2023	17/10/2023
					CT2750	Implantable cardiac electrical stimulators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	17/10/2023	17/10/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	17/10/2023	17/10/2023
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	17/10/2023	17/10/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	17/10/2023	17/10/2023
					CT2985	Active	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	17/10/2023	17/10/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	17/10/2023	17/10/2023
					CT3001	Implantable metal	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	17/10/2023	17/10/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	17/10/2023	17/10/2023
					CT3002	Implantable synthetic polymer	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	17/10/2023	17/10/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	17/10/2023	17/10/2023
					CT3006	Implantable inorganic material	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	17/10/2023	17/10/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	17/10/2023	17/10/2023
					CT314	Electrically powered	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	17/10/2023	17/10/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	17/10/2023	17/10/2023
					CT321	Long-term surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	17/10/2023	17/10/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	17/10/2023	17/10/2023
					CT449	Generators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	17/10/2023	17/10/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	17/10/2023	17/10/2023
					CT981	Single use	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	17/10/2023	17/10/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	17/10/2023	17/10/2023
					CT983	Surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	17/10/2023	17/10/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	17/10/2023	17/10/2023
14306	2023	All other reportable incidents	47265	Dual-chamber implantable pacemaker, rate-responsive	CT1685	Cardiac pacemakers	E0112	Dizziness	A071102	Failure to Interrogate	17/10/2023	17/10/2023
									A071204	No Pacing	17/10/2023	17/10/2023
							E0601	Arrhythmia	A071102	Failure to Interrogate	17/10/2023	17/10/2023
									A071204	No Pacing	17/10/2023	17/10/2023
							F1901	Additional Surgery	A071102	Failure to Interrogate	17/10/2023	17/10/2023
									A071204	No Pacing	17/10/2023	17/10/2023
					CT257	Implantable cardiac pacemakers	E0112	Dizziness	A071102	Failure to Interrogate	17/10/2023	17/10/2023
									A071204	No Pacing	17/10/2023	17/10/2023
							E0601	Arrhythmia	A071102	Failure to Interrogate	17/10/2023	17/10/2023
									A071204	No Pacing	17/10/2023	17/10/2023
							F1901	Additional Surgery	A071102	Failure to Interrogate	17/10/2023	17/10/2023
									A071204	No Pacing	17/10/2023	17/10/2023
					CT2866	Implantable cardiac electrical stimulators	E0112	Dizziness	A071102	Failure to Interrogate	17/10/2023	17/10/2023
									A071204	No Pacing	17/10/2023	17/10/2023
							E0601	Arrhythmia	A071102	Failure to Interrogate	17/10/2023	17/10/2023

									A071204	No Pacing	17/10/2023	17/10/2023
							F1901	Additional Surgery	A071102	Failure to Interrogate	17/10/2023	17/10/2023
									A071204	No Pacing	17/10/2023	17/10/2023
14307	2023	All other reportable incidents	47265	Dual-chamber implantable pacemaker, rate-responsive	CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	29/06/2023	29/06/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	29/06/2023	29/06/2023
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	29/06/2023	29/06/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	29/06/2023	29/06/2023
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	29/06/2023	29/06/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	29/06/2023	29/06/2023
14308	2023		47265	Dual-chamber implantable pacemaker, rate-responsive	CT141	Electrodes/Leads and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	17/10/2023	17/10/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	17/10/2023	17/10/2023
					CT145	Cardiovascular devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	17/10/2023	17/10/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	17/10/2023	17/10/2023
					CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	17/10/2023	17/10/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	17/10/2023	17/10/2023
					CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	17/10/2023	17/10/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	17/10/2023	17/10/2023
					CT233	Surgical	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	17/10/2023	17/10/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	17/10/2023	17/10/2023
					CT2406	Implantable	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	17/10/2023	17/10/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	17/10/2023	17/10/2023
					CT2407	Active implantable	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	17/10/2023	17/10/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	17/10/2023	17/10/2023
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	17/10/2023	17/10/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	17/10/2023	17/10/2023
					CT2727	Active implant devices/ancillaries	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	17/10/2023	17/10/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	17/10/2023	17/10/2023
					CT2750	Implantable cardiac electrical stimulators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	17/10/2023	17/10/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	17/10/2023	17/10/2023
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	17/10/2023	17/10/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	17/10/2023	17/10/2023
					CT2985	Active	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	17/10/2023	17/10/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	17/10/2023	17/10/2023
					CT3001	Implantable metal	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	17/10/2023	17/10/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	17/10/2023	17/10/2023
					CT3002	Implantable synthetic polymer	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	17/10/2023	17/10/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	17/10/2023	17/10/2023
					CT3006	Implantable inorganic material	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	17/10/2023	17/10/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	17/10/2023	17/10/2023
					CT314	Electrically powered	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	17/10/2023	17/10/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	17/10/2023	17/10/2023
					CT321	Long-term surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	17/10/2023	17/10/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	17/10/2023	17/10/2023
					CT449	Generators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	17/10/2023	17/10/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	17/10/2023	17/10/2023
					CT981	Single use	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	17/10/2023	17/10/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	17/10/2023	17/10/2023
					CT983	Surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	17/10/2023	17/10/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	17/10/2023	17/10/2023
14309	2023		47265	Dual-chamber implantable pacemaker, rate-responsive	CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	17/10/2023	17/10/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	17/10/2023	17/10/2023
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	17/10/2023	17/10/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	17/10/2023	17/10/2023
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	17/10/2023	17/10/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	17/10/2023	17/10/2023
14310	2023	All other reportable incidents	47265	Dual-chamber implantable pacemaker, rate-responsive	CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F1901	Additional Surgery	A070504	Premature Discharge of Battery		
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F1901	Additional Surgery	A070504	Premature Discharge of Battery		
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F1901	Additional Surgery	A070504	Premature Discharge of Battery		
14311	2023		47265	Dual-chamber implantable pacemaker, rate-responsive	CT141	Electrodes/Leads and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	11/10/2023	11/10/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	11/10/2023	11/10/2023
					CT145	Cardiovascular devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	11/10/2023	11/10/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	11/10/2023	11/10/2023
					CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	11/10/2023	11/10/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	11/10/2023	11/10/2023
					CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	11/10/2023	11/10/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	11/10/2023	11/10/2023
					CT233	Surgical	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	11/10/2023	11/10/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	11/10/2023	11/10/2023
					CT2406	Implantable	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	11/10/2023	11/10/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	11/10/2023	11/10/2023
					CT2407	Active implantable	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	11/10/2023	11/10/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	11/10/2023	11/10/2023
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	11/10/2023	11/10/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	11/10/2023	11/10/2023
					CT2727	Active implant devices/ancillaries	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	11/10/2023	11/10/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	11/10/2023	11/10/2023
					CT2750	Implantable cardiac electrical stimulators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	11/10/2023	11/10/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	11/10/2023	11/10/2023
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	11/10/2023	11/10/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	11/10/2023	11/10/2023
					CT2985	Active	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	11/10/2023	11/10/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	11/10/2023	11/10/2023
					CT3001	Implantable metal	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	11/10/2023	11/10/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	11/10/2023	11/10/2023
					CT3002	Implantable synthetic polymer	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	11/10/2023	11/10/2023

						F1901	Additional Surgery	A24	No Apparent Device Problem	11/10/2023	11/10/2023
				CT3006	Implantable inorganic material	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	11/10/2023	11/10/2023
						F1901	Additional Surgery	A24	No Apparent Device Problem	11/10/2023	11/10/2023
				CT314	Electrically powered	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	11/10/2023	11/10/2023
						F1901	Additional Surgery	A24	No Apparent Device Problem	11/10/2023	11/10/2023
				CT321	Long-term surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	11/10/2023	11/10/2023
						F1901	Additional Surgery	A24	No Apparent Device Problem	11/10/2023	11/10/2023
				CT449	Generators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	11/10/2023	11/10/2023
						F1901	Additional Surgery	A24	No Apparent Device Problem	11/10/2023	11/10/2023
				CT981	Single use	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	11/10/2023	11/10/2023
						F1901	Additional Surgery	A24	No Apparent Device Problem	11/10/2023	11/10/2023
				CT983	Surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	11/10/2023	11/10/2023
						F1901	Additional Surgery	A24	No Apparent Device Problem	11/10/2023	11/10/2023
14312	2023		47265	Dual-chamber implantable pacemaker, rate-responsive	CT141	Electrodes/Leads and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	11/10/2023	11/10/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	11/10/2023	11/10/2023
					CT145	Cardiovascular devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	11/10/2023	11/10/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	11/10/2023	11/10/2023
					CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	11/10/2023	11/10/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	11/10/2023	11/10/2023
					CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	11/10/2023	11/10/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	11/10/2023	11/10/2023
					CT233	Surgical	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	11/10/2023	11/10/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	11/10/2023	11/10/2023
					CT2406	Implantable	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	11/10/2023	11/10/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	11/10/2023	11/10/2023
					CT2407	Active implantable	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	11/10/2023	11/10/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	11/10/2023	11/10/2023
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	11/10/2023	11/10/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	11/10/2023	11/10/2023
					CT2727	Active implant devices/ancillaries	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	11/10/2023	11/10/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	11/10/2023	11/10/2023
					CT2750	Implantable cardiac electrical stimulators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	11/10/2023	11/10/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	11/10/2023	11/10/2023
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	11/10/2023	11/10/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	11/10/2023	11/10/2023
					CT2985	Active	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	11/10/2023	11/10/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	11/10/2023	11/10/2023
					CT3001	Implantable metal	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	11/10/2023	11/10/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	11/10/2023	11/10/2023
					CT3002	Implantable synthetic polymer	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	11/10/2023	11/10/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	11/10/2023	11/10/2023
					CT3006	Implantable inorganic material	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	11/10/2023	11/10/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	11/10/2023	11/10/2023
					CT314	Electrically powered	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	11/10/2023	11/10/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	11/10/2023	11/10/2023
					CT321	Long-term surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	11/10/2023	11/10/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	11/10/2023	11/10/2023
					CT449	Generators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	11/10/2023	11/10/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	11/10/2023	11/10/2023
					CT981	Single use	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	11/10/2023	11/10/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	11/10/2023	11/10/2023
					CT983	Surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	11/10/2023	11/10/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	11/10/2023	11/10/2023
14313	2023		47265	Dual-chamber implantable pacemaker, rate-responsive	CT141	Electrodes/Leads and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	03/10/2023	03/10/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	03/10/2023	03/10/2023
					CT145	Cardiovascular devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	03/10/2023	03/10/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	03/10/2023	03/10/2023
					CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	03/10/2023	03/10/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	03/10/2023	03/10/2023
					CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	03/10/2023	03/10/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	03/10/2023	03/10/2023
					CT233	Surgical	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	03/10/2023	03/10/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	03/10/2023	03/10/2023
					CT2406	Implantable	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	03/10/2023	03/10/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	03/10/2023	03/10/2023
					CT2407	Active implantable	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	03/10/2023	03/10/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	03/10/2023	03/10/2023
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	03/10/2023	03/10/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	03/10/2023	03/10/2023
					CT2727	Active implant devices/ancillaries	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	03/10/2023	03/10/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	03/10/2023	03/10/2023
					CT2750	Implantable cardiac electrical stimulators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	03/10/2023	03/10/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	03/10/2023	03/10/2023
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	03/10/2023	03/10/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	03/10/2023	03/10/2023
					CT2985	Active	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	03/10/2023	03/10/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	03/10/2023	03/10/2023
					CT3001	Implantable metal	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	03/10/2023	03/10/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	03/10/2023	03/10/2023
					CT3002	Implantable synthetic polymer	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	03/10/2023	03/10/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	03/10/2023	03/10/2023
					CT3006	Implantable inorganic material	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	03/10/2023	03/10/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	03/10/2023	03/10/2023
					CT314	Electrically powered	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	03/10/2023	03/10/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	03/10/2023	03/10/2023
					CT321	Long-term surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	03/10/2023	03/10/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	03/10/2023	03/10/2023
					CT449	Generators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	03/10/2023	03/10/2023

						F1901	Additional Surgery	A24	No Apparent Device Problem	03/10/2023	03/10/2023
				CT981	Single use	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	03/10/2023	03/10/2023
						F1901	Additional Surgery	A24	No Apparent Device Problem	03/10/2023	03/10/2023
				CT983	Surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	03/10/2023	03/10/2023
						F1901	Additional Surgery	A24	No Apparent Device Problem	03/10/2023	03/10/2023
14314	2023		47265	Dual-chamber implantable pacemaker, rate-responsive	CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	19/07/2023	19/07/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	19/07/2023	19/07/2023
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	19/07/2023	19/07/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	19/07/2023	19/07/2023
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	19/07/2023	19/07/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	19/07/2023	19/07/2023
14315	2023		47265	Dual-chamber implantable pacemaker, rate-responsive	CT141	Electrodes/Leads and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	26/09/2023	26/09/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	26/09/2023	26/09/2023
					CT145	Cardiovascular devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	26/09/2023	26/09/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	26/09/2023	26/09/2023
					CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	26/09/2023	26/09/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	26/09/2023	26/09/2023
					CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	26/09/2023	26/09/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	26/09/2023	26/09/2023
					CT233	Surgical	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	26/09/2023	26/09/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	26/09/2023	26/09/2023
					CT2406	Implantable	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	26/09/2023	26/09/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	26/09/2023	26/09/2023
					CT2407	Active implantable	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	26/09/2023	26/09/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	26/09/2023	26/09/2023
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	26/09/2023	26/09/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	26/09/2023	26/09/2023
					CT2727	Active implant devices/ancillaries	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	26/09/2023	26/09/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	26/09/2023	26/09/2023
					CT2750	Implantable cardiac electrical stimulators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	26/09/2023	26/09/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	26/09/2023	26/09/2023
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	26/09/2023	26/09/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	26/09/2023	26/09/2023
					CT2985	Active	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	26/09/2023	26/09/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	26/09/2023	26/09/2023
					CT3001	Implantable metal	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	26/09/2023	26/09/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	26/09/2023	26/09/2023
					CT3002	Implantable synthetic polymer	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	26/09/2023	26/09/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	26/09/2023	26/09/2023
					CT3006	Implantable inorganic material	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	26/09/2023	26/09/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	26/09/2023	26/09/2023
					CT314	Electrically powered	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	26/09/2023	26/09/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	26/09/2023	26/09/2023
					CT321	Long-term surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	26/09/2023	26/09/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	26/09/2023	26/09/2023
					CT449	Generators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	26/09/2023	26/09/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	26/09/2023	26/09/2023
					CT981	Single use	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	26/09/2023	26/09/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	26/09/2023	26/09/2023
					CT983	Surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	26/09/2023	26/09/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	26/09/2023	26/09/2023
14316	2023		47265	Dual-chamber implantable pacemaker, rate-responsive	CT141	Electrodes/Leads and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	10/10/2023	10/10/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	10/10/2023	10/10/2023
					CT145	Cardiovascular devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	10/10/2023	10/10/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	10/10/2023	10/10/2023
					CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	10/10/2023	10/10/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	10/10/2023	10/10/2023
					CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	10/10/2023	10/10/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	10/10/2023	10/10/2023
					CT233	Surgical	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	10/10/2023	10/10/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	10/10/2023	10/10/2023
					CT2406	Implantable	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	10/10/2023	10/10/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	10/10/2023	10/10/2023
					CT2407	Active implantable	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	10/10/2023	10/10/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	10/10/2023	10/10/2023
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	10/10/2023	10/10/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	10/10/2023	10/10/2023
					CT2727	Active implant devices/ancillaries	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	10/10/2023	10/10/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	10/10/2023	10/10/2023
					CT2750	Implantable cardiac electrical stimulators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	10/10/2023	10/10/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	10/10/2023	10/10/2023
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	10/10/2023	10/10/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	10/10/2023	10/10/2023
					CT2985	Active	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	10/10/2023	10/10/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	10/10/2023	10/10/2023
					CT3001	Implantable metal	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	10/10/2023	10/10/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	10/10/2023	10/10/2023
					CT3002	Implantable synthetic polymer	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	10/10/2023	10/10/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	10/10/2023	10/10/2023
					CT3006	Implantable inorganic material	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	10/10/2023	10/10/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	10/10/2023	10/10/2023
					CT314	Electrically powered	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	10/10/2023	10/10/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	10/10/2023	10/10/2023
					CT321	Long-term surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	10/10/2023	10/10/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	10/10/2023	10/10/2023
					CT449	Generators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	10/10/2023	10/10/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	10/10/2023	10/10/2023
					CT981	Single use	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	10/10/2023	10/10/2023

							F1901	Additional Surgery	A24	No Apparent Device Problem	10/10/2023	10/10/2023
					CT983	Surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	10/10/2023	10/10/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	10/10/2023	10/10/2023
14317	2023	All other reportable incidents	17882	Physiologic-monitoring defibrillation system	CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070803	Failure to Power Up		
							F27	Problem Identified During Non-Clinical Procedure	A070803	Failure to Power Up		
14318	2023	All other reportable incidents	47265	Dual-chamber implantable pacemaker, rate-responsive	CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	12/10/2023	12/10/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	12/10/2023	12/10/2023
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	12/10/2023	12/10/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	12/10/2023	12/10/2023
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	12/10/2023	12/10/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	12/10/2023	12/10/2023
14319	2023		47265	Dual-chamber implantable pacemaker, rate-responsive	CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	09/10/2023	09/10/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	09/10/2023	09/10/2023
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	09/10/2023	09/10/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	09/10/2023	09/10/2023
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	09/10/2023	09/10/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	09/10/2023	09/10/2023
14320	2023	All other reportable incidents	35223	Endocardial/interventricular septal pacing lead	CT1684	Cardiac pacemakers and associated devices	E0601	Arrhythmia	A070910	Undersensing		
									A0712	Pacing Problem		
							F12	Serious Injury/ Illness/ Impairment	A070910	Undersensing		
									A0712	Pacing Problem		
14321	2023	All other reportable incidents	37265	Dual-chamber implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E0601	Arrhythmia	A071206	Pacing Asynchronously		
									A071302	Inappropriate/Inadequate Shock/Stimulation		
							F12	Serious Injury/ Illness/ Impairment	A071206	Pacing Asynchronously		
									A071302	Inappropriate/Inadequate Shock/Stimulation		
14322	2023	All other reportable incidents	47265	Dual-chamber implantable pacemaker, rate-responsive	CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem	08/06/2023	08/06/2023
									A070504	Premature Discharge of Battery	08/06/2023	08/06/2023
							F08	Hospitalization or Prolonged Hospitalization	A0705	Battery Problem	08/06/2023	08/06/2023
									A070504	Premature Discharge of Battery	08/06/2023	08/06/2023
							F1905	Device Revision or Replacement	A0705	Battery Problem	08/06/2023	08/06/2023
									A070504	Premature Discharge of Battery	08/06/2023	08/06/2023
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem	08/06/2023	08/06/2023
									A070504	Premature Discharge of Battery	08/06/2023	08/06/2023
							F08	Hospitalization or Prolonged Hospitalization	A0705	Battery Problem	08/06/2023	08/06/2023
									A070504	Premature Discharge of Battery	08/06/2023	08/06/2023
							F1905	Device Revision or Replacement	A0705	Battery Problem	08/06/2023	08/06/2023
									A070504	Premature Discharge of Battery	08/06/2023	08/06/2023
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem	08/06/2023	08/06/2023
									A070504	Premature Discharge of Battery	08/06/2023	08/06/2023
							F08	Hospitalization or Prolonged Hospitalization	A0705	Battery Problem	08/06/2023	08/06/2023
									A070504	Premature Discharge of Battery	08/06/2023	08/06/2023
							F1905	Device Revision or Replacement	A0705	Battery Problem	08/06/2023	08/06/2023
									A070504	Premature Discharge of Battery	08/06/2023	08/06/2023
14323	2023	All other reportable incidents	63188	Left-ventricular wireless pacing system sensor/transmitter	CT141	Electrodes/Leads and associated devices			A0708	Power Problem		
					CT145	Cardiovascular devices			A0708	Power Problem		
					CT1577	Wireless monitoring system transmitters			A0708	Power Problem		
					CT1684	Cardiac pacemakers and associated devices			A0708	Power Problem		
					CT2317	Transmitters/Receivers/Transceivers			A0708	Power Problem		
					CT233	Surgical			A0708	Power Problem		
					CT2406	Implantable			A0708	Power Problem		
					CT2407	Active implantable			A0708	Power Problem		
					CT255	Transducers and associated devices			A0708	Power Problem		
					CT2726	Physiological monitoring/measuring devices			A0708	Power Problem		
					CT2727	Active implant devices/ancillaries			A0708	Power Problem		
					CT2750	Implantable cardiac electrical stimulators and associated devices			A0708	Power Problem		
					CT2885	Physiological monitoring/measuring system transmitters/receivers/transceivers			A0708	Power Problem		
					CT2985	Active			A0708	Power Problem		
					CT3001	Implantable metal			A0708	Power Problem		
					CT3002	Implantable synthetic polymer			A0708	Power Problem		
					CT3006	Implantable inorganic material			A0708	Power Problem		
					CT314	Electrically powered			A0708	Power Problem		
					CT321	Long-term surgical invasive			A0708	Power Problem		
					CT384	Sensors and associated devices			A0708	Power Problem		
					CT449	Generators and associated devices			A0708	Power Problem		
					CT981	Single use			A0708	Power Problem		
					CT983	Surgical invasive			A0708	Power Problem		
					CT994	Sensors			A0708	Power Problem		
14324	2023	All other reportable incidents	35223	Endocardial/interventricular septal pacing lead	CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070909	Oversensing		
							F26	No Health Consequences or Impact	A070909	Oversensing		
14325	2023	All other reportable incidents	35822	External pacemaker, transcutaneous pacing	CT1685	Cardiac pacemakers	E2401	Insufficient Information	A070101	Failure to Capture		
							F24	Insufficient Health Impact Information	A070101	Failure to Capture		
14326	2023	All other reportable incidents	47263	Cardiac resynchronization therapy implantable pacemaker	CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A0908	Incorrect, Inadequate or Imprecise Result or Readings	29/11/2022	29/11/2022
									A160503	Inappropriate or Unexpected Reset	29/11/2022	29/11/2022
							F08	Hospitalization or Prolonged Hospitalization	A0908	Incorrect, Inadequate or Imprecise Result or Readings	29/11/2022	29/11/2022
									A160503	Inappropriate or Unexpected Reset	29/11/2022	29/11/2022
							F1905	Device Revision or Replacement	A0908	Incorrect, Inadequate or Imprecise Result or Readings	29/11/2022	29/11/2022
									A160503	Inappropriate or Unexpected Reset	29/11/2022	29/11/2022
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A0908	Incorrect, Inadequate or Imprecise Result or Readings	29/11/2022	29/11/2022
									A160503	Inappropriate or Unexpected Reset	29/11/2022	29/11/2022
							F08	Hospitalization or Prolonged Hospitalization	A0908	Incorrect, Inadequate or Imprecise Result or Readings	29/11/2022	29/11/2022
									A160503	Inappropriate or Unexpected Reset	29/11/2022	29/11/2022
							F1905	Device Revision or Replacement	A0908	Incorrect, Inadequate or Imprecise Result or Readings	29/11/2022	29/11/2022

									A160503	Inappropriate or Unexpected Reset	29/11/2022	29/11/2022
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A0908	Incorrect, Inadequate or Imprecise Result or Readings	29/11/2022	29/11/2022
									A160503	Inappropriate or Unexpected Reset	29/11/2022	29/11/2022
							F08	Hospitalization or Prolonged Hospitalization	A0908	Incorrect, Inadequate or Imprecise Result or Readings	29/11/2022	29/11/2022
									A160503	Inappropriate or Unexpected Reset	29/11/2022	29/11/2022
							F1905	Device Revision or Replacement	A0908	Incorrect, Inadequate or Imprecise Result or Readings	29/11/2022	29/11/2022
									A160503	Inappropriate or Unexpected Reset	29/11/2022	29/11/2022
14327	2023	All other reportable incidents	47265	Dual-chamber implantable pacemaker, rate-responsive	CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
14328	2023	All other reportable incidents	47265	Dual-chamber implantable pacemaker, rate-responsive	CT141	Electrodes/Leads and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem		
							F26	No Health Consequences or Impact	A0705	Battery Problem		
					CT145	Cardiovascular devices	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem		
							F26	No Health Consequences or Impact	A0705	Battery Problem		
					CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem		
							F26	No Health Consequences or Impact	A0705	Battery Problem		
					CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem		
							F26	No Health Consequences or Impact	A0705	Battery Problem		
					CT233	Surgical	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem		
							F26	No Health Consequences or Impact	A0705	Battery Problem		
					CT2406	Implantable	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem		
							F26	No Health Consequences or Impact	A0705	Battery Problem		
					CT2407	Active implantable	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem		
							F26	No Health Consequences or Impact	A0705	Battery Problem		
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem		
							F26	No Health Consequences or Impact	A0705	Battery Problem		
					CT2727	Active implant devices/ancillaries	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem		
							F26	No Health Consequences or Impact	A0705	Battery Problem		
					CT2750	Implantable cardiac electrical stimulators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem		
							F26	No Health Consequences or Impact	A0705	Battery Problem		
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem		
							F26	No Health Consequences or Impact	A0705	Battery Problem		
					CT2985	Active	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem		
							F26	No Health Consequences or Impact	A0705	Battery Problem		
					CT3001	Implantable metal	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem		
							F26	No Health Consequences or Impact	A0705	Battery Problem		
					CT3002	Implantable synthetic polymer	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem		
							F26	No Health Consequences or Impact	A0705	Battery Problem		
					CT3006	Implantable inorganic material	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem		
							F26	No Health Consequences or Impact	A0705	Battery Problem		
					CT314	Electrically powered	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem		
							F26	No Health Consequences or Impact	A0705	Battery Problem		
					CT321	Long-term surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem		
							F26	No Health Consequences or Impact	A0705	Battery Problem		
					CT449	Generators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem		
							F26	No Health Consequences or Impact	A0705	Battery Problem		
					CT981	Single use	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem		
							F26	No Health Consequences or Impact	A0705	Battery Problem		
					CT983	Surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem		
							F26	No Health Consequences or Impact	A0705	Battery Problem		
14329	2023	All other reportable incidents	47265	Dual-chamber implantable pacemaker, rate-responsive	CT141	Electrodes/Leads and associated devices	E0112	Dizziness	A070101	Failure to Capture	19/10/2023	19/10/2023
									A0712	Pacing Problem	19/10/2023	19/10/2023
									A0908	Incorrect, Inadequate or Imprecise Result or Readings	19/10/2023	19/10/2023
							E2103	Device Overstimulation of Tissue	A070101	Failure to Capture	19/10/2023	19/10/2023
									A0712	Pacing Problem	19/10/2023	19/10/2023
									A0908	Incorrect, Inadequate or Imprecise Result or Readings	19/10/2023	19/10/2023
							F08	Hospitalization or Prolonged Hospitalization	A070101	Failure to Capture	19/10/2023	19/10/2023
									A0712	Pacing Problem	19/10/2023	19/10/2023
									A0908	Incorrect, Inadequate or Imprecise Result or Readings	19/10/2023	19/10/2023
							F1905	Device Revision or Replacement	A070101	Failure to Capture	19/10/2023	19/10/2023
									A0712	Pacing Problem	19/10/2023	19/10/2023
									A0908	Incorrect, Inadequate or Imprecise Result or Readings	19/10/2023	19/10/2023
					CT145	Cardiovascular devices	E0112	Dizziness	A070101	Failure to Capture	19/10/2023	19/10/2023
									A0712	Pacing Problem	19/10/2023	19/10/2023
									A0908	Incorrect, Inadequate or Imprecise Result or Readings	19/10/2023	19/10/2023
							E2103	Device Overstimulation of Tissue	A070101	Failure to Capture	19/10/2023	19/10/2023
									A0712	Pacing Problem	19/10/2023	19/10/2023
									A0908	Incorrect, Inadequate or Imprecise Result or Readings	19/10/2023	19/10/2023
							F08	Hospitalization or Prolonged Hospitalization	A070101	Failure to Capture	19/10/2023	19/10/2023
									A0712	Pacing Problem	19/10/2023	19/10/2023
									A0908	Incorrect, Inadequate or Imprecise Result or Readings	19/10/2023	19/10/2023

							F1905	Device Revision or Replacement	A070101	Failure to Capture	19/10/2023	19/10/2023
									A0712	Pacing Problem	19/10/2023	19/10/2023
									A0908	Incorrect, Inadequate or Imprecise Result or Readings	19/10/2023	19/10/2023
					CT1684	Cardiac pacemakers and associated devices	E0112	Dizziness	A070101	Failure to Capture	19/10/2023	19/10/2023
									A0712	Pacing Problem	19/10/2023	19/10/2023
									A0908	Incorrect, Inadequate or Imprecise Result or Readings	19/10/2023	19/10/2023
							E2103	Device Overstimulation of Tissue	A070101	Failure to Capture	19/10/2023	19/10/2023
									A0712	Pacing Problem	19/10/2023	19/10/2023
									A0908	Incorrect, Inadequate or Imprecise Result or Readings	19/10/2023	19/10/2023
							F08	Hospitalization or Prolonged Hospitalization	A070101	Failure to Capture	19/10/2023	19/10/2023
									A0712	Pacing Problem	19/10/2023	19/10/2023
									A0908	Incorrect, Inadequate or Imprecise Result or Readings	19/10/2023	19/10/2023
							F1905	Device Revision or Replacement	A070101	Failure to Capture	19/10/2023	19/10/2023
									A0712	Pacing Problem	19/10/2023	19/10/2023
									A0908	Incorrect, Inadequate or Imprecise Result or Readings	19/10/2023	19/10/2023
					CT1685	Cardiac pacemakers	E0112	Dizziness	A070101	Failure to Capture	19/10/2023	19/10/2023
									A0712	Pacing Problem	19/10/2023	19/10/2023
									A0908	Incorrect, Inadequate or Imprecise Result or Readings	19/10/2023	19/10/2023
							E2103	Device Overstimulation of Tissue	A070101	Failure to Capture	19/10/2023	19/10/2023
									A0712	Pacing Problem	19/10/2023	19/10/2023
									A0908	Incorrect, Inadequate or Imprecise Result or Readings	19/10/2023	19/10/2023
							F08	Hospitalization or Prolonged Hospitalization	A070101	Failure to Capture	19/10/2023	19/10/2023
									A0712	Pacing Problem	19/10/2023	19/10/2023
									A0908	Incorrect, Inadequate or Imprecise Result or Readings	19/10/2023	19/10/2023
							F1905	Device Revision or Replacement	A070101	Failure to Capture	19/10/2023	19/10/2023
									A0712	Pacing Problem	19/10/2023	19/10/2023
									A0908	Incorrect, Inadequate or Imprecise Result or Readings	19/10/2023	19/10/2023
					CT233	Surgical	E0112	Dizziness	A070101	Failure to Capture	19/10/2023	19/10/2023
									A0712	Pacing Problem	19/10/2023	19/10/2023
									A0908	Incorrect, Inadequate or Imprecise Result or Readings	19/10/2023	19/10/2023
							E2103	Device Overstimulation of Tissue	A070101	Failure to Capture	19/10/2023	19/10/2023
									A0712	Pacing Problem	19/10/2023	19/10/2023
									A0908	Incorrect, Inadequate or Imprecise Result or Readings	19/10/2023	19/10/2023
							F08	Hospitalization or Prolonged Hospitalization	A070101	Failure to Capture	19/10/2023	19/10/2023
									A0712	Pacing Problem	19/10/2023	19/10/2023
									A0908	Incorrect, Inadequate or Imprecise Result or Readings	19/10/2023	19/10/2023
							F1905	Device Revision or Replacement	A070101	Failure to Capture	19/10/2023	19/10/2023
									A0712	Pacing Problem	19/10/2023	19/10/2023
									A0908	Incorrect, Inadequate or Imprecise Result or Readings	19/10/2023	19/10/2023
					CT2406	Implantable	E0112	Dizziness	A070101	Failure to Capture	19/10/2023	19/10/2023
									A0712	Pacing Problem	19/10/2023	19/10/2023
									A0908	Incorrect, Inadequate or Imprecise Result or Readings	19/10/2023	19/10/2023
							E2103	Device Overstimulation of Tissue	A070101	Failure to Capture	19/10/2023	19/10/2023
									A0712	Pacing Problem	19/10/2023	19/10/2023
									A0908	Incorrect, Inadequate or Imprecise Result or Readings	19/10/2023	19/10/2023
							F08	Hospitalization or Prolonged Hospitalization	A070101	Failure to Capture	19/10/2023	19/10/2023
									A0712	Pacing Problem	19/10/2023	19/10/2023
									A0908	Incorrect, Inadequate or Imprecise Result or Readings	19/10/2023	19/10/2023
							F1905	Device Revision or Replacement	A070101	Failure to Capture	19/10/2023	19/10/2023
									A0712	Pacing Problem	19/10/2023	19/10/2023
									A0908	Incorrect, Inadequate or Imprecise Result or Readings	19/10/2023	19/10/2023
					CT2407	Active implantable	E0112	Dizziness	A070101	Failure to Capture	19/10/2023	19/10/2023
									A0712	Pacing Problem	19/10/2023	19/10/2023
									A0908	Incorrect, Inadequate or Imprecise Result or Readings	19/10/2023	19/10/2023
							E2103	Device Overstimulation of Tissue	A070101	Failure to Capture	19/10/2023	19/10/2023
									A0712	Pacing Problem	19/10/2023	19/10/2023
									A0908	Incorrect, Inadequate or Imprecise Result or Readings	19/10/2023	19/10/2023
							F08	Hospitalization or Prolonged Hospitalization	A070101	Failure to Capture	19/10/2023	19/10/2023
									A0712	Pacing Problem	19/10/2023	19/10/2023
									A0908	Incorrect, Inadequate or Imprecise Result or Readings	19/10/2023	19/10/2023
							F1905	Device Revision or Replacement	A070101	Failure to Capture	19/10/2023	19/10/2023
									A0712	Pacing Problem	19/10/2023	19/10/2023
									A0908	Incorrect, Inadequate or Imprecise Result or Readings	19/10/2023	19/10/2023
					CT257	Implantable cardiac pacemakers	E0112	Dizziness	A070101	Failure to Capture	19/10/2023	19/10/2023
									A0712	Pacing Problem	19/10/2023	19/10/2023
									A0908	Incorrect, Inadequate or Imprecise Result or Readings	19/10/2023	19/10/2023
							E2103	Device Overstimulation of Tissue	A070101	Failure to Capture	19/10/2023	19/10/2023
									A0712	Pacing Problem	19/10/2023	19/10/2023
									A0908	Incorrect, Inadequate or Imprecise Result or Readings	19/10/2023	19/10/2023
							F08	Hospitalization or Prolonged Hospitalization	A070101	Failure to Capture	19/10/2023	19/10/2023

								A0712	Pacing Problem	19/10/2023	19/10/2023
								A0908	Incorrect, Inadequate or Imprecise Result or Readings	19/10/2023	19/10/2023
						F1905	Device Revision or Replacement	A070101	Failure to Capture	19/10/2023	19/10/2023
								A0712	Pacing Problem	19/10/2023	19/10/2023
								A0908	Incorrect, Inadequate or Imprecise Result or Readings	19/10/2023	19/10/2023
					CT2727	Active implant devices/ancillaries	E0112	Dizziness	A070101	Failure to Capture	19/10/2023	19/10/2023
									A0712	Pacing Problem	19/10/2023	19/10/2023
									A0908	Incorrect, Inadequate or Imprecise Result or Readings	19/10/2023	19/10/2023
							E2103	Device Overstimulation of Tissue	A070101	Failure to Capture	19/10/2023	19/10/2023
									A0712	Pacing Problem	19/10/2023	19/10/2023
									A0908	Incorrect, Inadequate or Imprecise Result or Readings	19/10/2023	19/10/2023
							F08	Hospitalization or Prolonged Hospitalization	A070101	Failure to Capture	19/10/2023	19/10/2023
									A0712	Pacing Problem	19/10/2023	19/10/2023
									A0908	Incorrect, Inadequate or Imprecise Result or Readings	19/10/2023	19/10/2023
							F1905	Device Revision or Replacement	A070101	Failure to Capture	19/10/2023	19/10/2023
									A0712	Pacing Problem	19/10/2023	19/10/2023
									A0908	Incorrect, Inadequate or Imprecise Result or Readings	19/10/2023	19/10/2023
					CT2750	Implantable cardiac electrical stimulators and associated devices	E0112	Dizziness	A070101	Failure to Capture	19/10/2023	19/10/2023
									A0712	Pacing Problem	19/10/2023	19/10/2023
									A0908	Incorrect, Inadequate or Imprecise Result or Readings	19/10/2023	19/10/2023
							E2103	Device Overstimulation of Tissue	A070101	Failure to Capture	19/10/2023	19/10/2023
									A0712	Pacing Problem	19/10/2023	19/10/2023
									A0908	Incorrect, Inadequate or Imprecise Result or Readings	19/10/2023	19/10/2023
							F08	Hospitalization or Prolonged Hospitalization	A070101	Failure to Capture	19/10/2023	19/10/2023
									A0712	Pacing Problem	19/10/2023	19/10/2023
									A0908	Incorrect, Inadequate or Imprecise Result or Readings	19/10/2023	19/10/2023
							F1905	Device Revision or Replacement	A070101	Failure to Capture	19/10/2023	19/10/2023
									A0712	Pacing Problem	19/10/2023	19/10/2023
									A0908	Incorrect, Inadequate or Imprecise Result or Readings	19/10/2023	19/10/2023
					CT2866	Implantable cardiac electrical stimulators	E0112	Dizziness	A070101	Failure to Capture	19/10/2023	19/10/2023
									A0712	Pacing Problem	19/10/2023	19/10/2023
									A0908	Incorrect, Inadequate or Imprecise Result or Readings	19/10/2023	19/10/2023
							E2103	Device Overstimulation of Tissue	A070101	Failure to Capture	19/10/2023	19/10/2023
									A0712	Pacing Problem	19/10/2023	19/10/2023
									A0908	Incorrect, Inadequate or Imprecise Result or Readings	19/10/2023	19/10/2023
							F08	Hospitalization or Prolonged Hospitalization	A070101	Failure to Capture	19/10/2023	19/10/2023
									A0712	Pacing Problem	19/10/2023	19/10/2023
									A0908	Incorrect, Inadequate or Imprecise Result or Readings	19/10/2023	19/10/2023
							F1905	Device Revision or Replacement	A070101	Failure to Capture	19/10/2023	19/10/2023
									A0712	Pacing Problem	19/10/2023	19/10/2023
									A0908	Incorrect, Inadequate or Imprecise Result or Readings	19/10/2023	19/10/2023
					CT2985	Active	E0112	Dizziness	A070101	Failure to Capture	19/10/2023	19/10/2023
									A0712	Pacing Problem	19/10/2023	19/10/2023
									A0908	Incorrect, Inadequate or Imprecise Result or Readings	19/10/2023	19/10/2023
							E2103	Device Overstimulation of Tissue	A070101	Failure to Capture	19/10/2023	19/10/2023
									A0712	Pacing Problem	19/10/2023	19/10/2023
									A0908	Incorrect, Inadequate or Imprecise Result or Readings	19/10/2023	19/10/2023
							F08	Hospitalization or Prolonged Hospitalization	A070101	Failure to Capture	19/10/2023	19/10/2023
									A0712	Pacing Problem	19/10/2023	19/10/2023
									A0908	Incorrect, Inadequate or Imprecise Result or Readings	19/10/2023	19/10/2023
							F1905	Device Revision or Replacement	A070101	Failure to Capture	19/10/2023	19/10/2023
									A0712	Pacing Problem	19/10/2023	19/10/2023
									A0908	Incorrect, Inadequate or Imprecise Result or Readings	19/10/2023	19/10/2023
					CT3001	Implantable metal	E0112	Dizziness	A070101	Failure to Capture	19/10/2023	19/10/2023
									A0712	Pacing Problem	19/10/2023	19/10/2023
									A0908	Incorrect, Inadequate or Imprecise Result or Readings	19/10/2023	19/10/2023
							E2103	Device Overstimulation of Tissue	A070101	Failure to Capture	19/10/2023	19/10/2023
									A0712	Pacing Problem	19/10/2023	19/10/2023
									A0908	Incorrect, Inadequate or Imprecise Result or Readings	19/10/2023	19/10/2023
							F08	Hospitalization or Prolonged Hospitalization	A070101	Failure to Capture	19/10/2023	19/10/2023
									A0712	Pacing Problem	19/10/2023	19/10/2023
									A0908	Incorrect, Inadequate or Imprecise Result or Readings	19/10/2023	19/10/2023
							F1905	Device Revision or Replacement	A070101	Failure to Capture	19/10/2023	19/10/2023
									A0712	Pacing Problem	19/10/2023	19/10/2023
									A0908	Incorrect, Inadequate or Imprecise Result or Readings	19/10/2023	19/10/2023
					CT3002	Implantable synthetic polymer	E0112	Dizziness	A070101	Failure to Capture	19/10/2023	19/10/2023
									A0712	Pacing Problem	19/10/2023	19/10/2023
									A0908	Incorrect, Inadequate or Imprecise Result or Readings	19/10/2023	19/10/2023
							E2103	Device Overstimulation of Tissue	A070101	Failure to Capture	19/10/2023	19/10/2023
									A0712	Pacing Problem	19/10/2023	19/10/2023

								A0908	Incorrect, Inadequate or Imprecise Result or Readings	19/10/2023	19/10/2023
						F08	Hospitalization or Prolonged Hospitalization	A070101	Failure to Capture	19/10/2023	19/10/2023
								A0712	Pacing Problem	19/10/2023	19/10/2023
								A0908	Incorrect, Inadequate or Imprecise Result or Readings	19/10/2023	19/10/2023
						F1905	Device Revision or Replacement	A070101	Failure to Capture	19/10/2023	19/10/2023
								A0712	Pacing Problem	19/10/2023	19/10/2023
								A0908	Incorrect, Inadequate or Imprecise Result or Readings	19/10/2023	19/10/2023
					CT3006	Implantable inorganic material	E0112	Dizziness	A070101	Failure to Capture	19/10/2023	19/10/2023
									A0712	Pacing Problem	19/10/2023	19/10/2023
									A0908	Incorrect, Inadequate or Imprecise Result or Readings	19/10/2023	19/10/2023
							E2103	Device Overstimulation of Tissue	A070101	Failure to Capture	19/10/2023	19/10/2023
									A0712	Pacing Problem	19/10/2023	19/10/2023
									A0908	Incorrect, Inadequate or Imprecise Result or Readings	19/10/2023	19/10/2023
							F08	Hospitalization or Prolonged Hospitalization	A070101	Failure to Capture	19/10/2023	19/10/2023
									A0712	Pacing Problem	19/10/2023	19/10/2023
									A0908	Incorrect, Inadequate or Imprecise Result or Readings	19/10/2023	19/10/2023
							F1905	Device Revision or Replacement	A070101	Failure to Capture	19/10/2023	19/10/2023
									A0712	Pacing Problem	19/10/2023	19/10/2023
									A0908	Incorrect, Inadequate or Imprecise Result or Readings	19/10/2023	19/10/2023
					CT314	Electrically powered	E0112	Dizziness	A070101	Failure to Capture	19/10/2023	19/10/2023
									A0712	Pacing Problem	19/10/2023	19/10/2023
									A0908	Incorrect, Inadequate or Imprecise Result or Readings	19/10/2023	19/10/2023
							E2103	Device Overstimulation of Tissue	A070101	Failure to Capture	19/10/2023	19/10/2023
									A0712	Pacing Problem	19/10/2023	19/10/2023
									A0908	Incorrect, Inadequate or Imprecise Result or Readings	19/10/2023	19/10/2023
							F08	Hospitalization or Prolonged Hospitalization	A070101	Failure to Capture	19/10/2023	19/10/2023
									A0712	Pacing Problem	19/10/2023	19/10/2023
									A0908	Incorrect, Inadequate or Imprecise Result or Readings	19/10/2023	19/10/2023
							F1905	Device Revision or Replacement	A070101	Failure to Capture	19/10/2023	19/10/2023
									A0712	Pacing Problem	19/10/2023	19/10/2023
									A0908	Incorrect, Inadequate or Imprecise Result or Readings	19/10/2023	19/10/2023
					CT321	Long-term surgical invasive	E0112	Dizziness	A070101	Failure to Capture	19/10/2023	19/10/2023
									A0712	Pacing Problem	19/10/2023	19/10/2023
									A0908	Incorrect, Inadequate or Imprecise Result or Readings	19/10/2023	19/10/2023
							E2103	Device Overstimulation of Tissue	A070101	Failure to Capture	19/10/2023	19/10/2023
									A0712	Pacing Problem	19/10/2023	19/10/2023
									A0908	Incorrect, Inadequate or Imprecise Result or Readings	19/10/2023	19/10/2023
							F08	Hospitalization or Prolonged Hospitalization	A070101	Failure to Capture	19/10/2023	19/10/2023
									A0712	Pacing Problem	19/10/2023	19/10/2023
									A0908	Incorrect, Inadequate or Imprecise Result or Readings	19/10/2023	19/10/2023
							F1905	Device Revision or Replacement	A070101	Failure to Capture	19/10/2023	19/10/2023
									A0712	Pacing Problem	19/10/2023	19/10/2023
									A0908	Incorrect, Inadequate or Imprecise Result or Readings	19/10/2023	19/10/2023
					CT449	Generators and associated devices	E0112	Dizziness	A070101	Failure to Capture	19/10/2023	19/10/2023
									A0712	Pacing Problem	19/10/2023	19/10/2023
									A0908	Incorrect, Inadequate or Imprecise Result or Readings	19/10/2023	19/10/2023
							E2103	Device Overstimulation of Tissue	A070101	Failure to Capture	19/10/2023	19/10/2023
									A0712	Pacing Problem	19/10/2023	19/10/2023
									A0908	Incorrect, Inadequate or Imprecise Result or Readings	19/10/2023	19/10/2023
							F08	Hospitalization or Prolonged Hospitalization	A070101	Failure to Capture	19/10/2023	19/10/2023
									A0712	Pacing Problem	19/10/2023	19/10/2023
									A0908	Incorrect, Inadequate or Imprecise Result or Readings	19/10/2023	19/10/2023
							F1905	Device Revision or Replacement	A070101	Failure to Capture	19/10/2023	19/10/2023
									A0712	Pacing Problem	19/10/2023	19/10/2023
									A0908	Incorrect, Inadequate or Imprecise Result or Readings	19/10/2023	19/10/2023
					CT981	Single use	E0112	Dizziness	A070101	Failure to Capture	19/10/2023	19/10/2023
									A0712	Pacing Problem	19/10/2023	19/10/2023
									A0908	Incorrect, Inadequate or Imprecise Result or Readings	19/10/2023	19/10/2023
							E2103	Device Overstimulation of Tissue	A070101	Failure to Capture	19/10/2023	19/10/2023
									A0712	Pacing Problem	19/10/2023	19/10/2023
									A0908	Incorrect, Inadequate or Imprecise Result or Readings	19/10/2023	19/10/2023
							F08	Hospitalization or Prolonged Hospitalization	A070101	Failure to Capture	19/10/2023	19/10/2023
									A0712	Pacing Problem	19/10/2023	19/10/2023
									A0908	Incorrect, Inadequate or Imprecise Result or Readings	19/10/2023	19/10/2023
							F1905	Device Revision or Replacement	A070101	Failure to Capture	19/10/2023	19/10/2023
									A0712	Pacing Problem	19/10/2023	19/10/2023
									A0908	Incorrect, Inadequate or Imprecise Result or Readings	19/10/2023	19/10/2023
					CT983	Surgical invasive	E0112	Dizziness	A070101	Failure to Capture	19/10/2023	19/10/2023
									A0712	Pacing Problem	19/10/2023	19/10/2023

								A0908	Incorrect, Inadequate or Imprecise Result or Readings	19/10/2023	19/10/2023
						E2103	Device Overstimulation of Tissue	A070101	Failure to Capture	19/10/2023	19/10/2023
								A0712	Pacing Problem	19/10/2023	19/10/2023
								A0908	Incorrect, Inadequate or Imprecise Result or Readings	19/10/2023	19/10/2023
						F08	Hospitalization or Prolonged Hospitalization	A070101	Failure to Capture	19/10/2023	19/10/2023
								A0712	Pacing Problem	19/10/2023	19/10/2023
								A0908	Incorrect, Inadequate or Imprecise Result or Readings	19/10/2023	19/10/2023
						F1905	Device Revision or Replacement	A070101	Failure to Capture	19/10/2023	19/10/2023
								A0712	Pacing Problem	19/10/2023	19/10/2023
								A0908	Incorrect, Inadequate or Imprecise Result or Readings	19/10/2023	19/10/2023
14330	2023	All other reportable incidents	37265	Dual-chamber implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A090811	Unable to Obtain Readings		
							F26	No Health Consequences or Impact	A090811	Unable to Obtain Readings		
14331	2023	All other reportable incidents	17882	Physiologic-monitoring defibrillation system	CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070803	Failure to Power Up		
							F27	Problem Identified During Non-Clinical Procedure	A070803	Failure to Power Up		
14332	2023	All other reportable incidents	47265	Dual-chamber implantable pacemaker, rate-responsive	CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A070909	Oversensing	02/07/2023	02/07/2023
									A0712	Pacing Problem	02/07/2023	02/07/2023
									A0908	Incorrect, Inadequate or Imprecise Result or Readings	02/07/2023	02/07/2023
							F08	Hospitalization or Prolonged Hospitalization	A070909	Oversensing	02/07/2023	02/07/2023
									A0712	Pacing Problem	02/07/2023	02/07/2023
									A0908	Incorrect, Inadequate or Imprecise Result or Readings	02/07/2023	02/07/2023
							F1001	Absence of Treatment	A070909	Oversensing	02/07/2023	02/07/2023
									A0712	Pacing Problem	02/07/2023	02/07/2023
									A0908	Incorrect, Inadequate or Imprecise Result or Readings	02/07/2023	02/07/2023
							F1905	Device Revision or Replacement	A070909	Oversensing	02/07/2023	02/07/2023
									A0712	Pacing Problem	02/07/2023	02/07/2023
									A0908	Incorrect, Inadequate or Imprecise Result or Readings	02/07/2023	02/07/2023
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A070909	Oversensing	02/07/2023	02/07/2023
									A0712	Pacing Problem	02/07/2023	02/07/2023
									A0908	Incorrect, Inadequate or Imprecise Result or Readings	02/07/2023	02/07/2023
							F08	Hospitalization or Prolonged Hospitalization	A070909	Oversensing	02/07/2023	02/07/2023
									A0712	Pacing Problem	02/07/2023	02/07/2023
									A0908	Incorrect, Inadequate or Imprecise Result or Readings	02/07/2023	02/07/2023
							F1001	Absence of Treatment	A070909	Oversensing	02/07/2023	02/07/2023
									A0712	Pacing Problem	02/07/2023	02/07/2023
									A0908	Incorrect, Inadequate or Imprecise Result or Readings	02/07/2023	02/07/2023
							F1905	Device Revision or Replacement	A070909	Oversensing	02/07/2023	02/07/2023
									A0712	Pacing Problem	02/07/2023	02/07/2023
									A0908	Incorrect, Inadequate or Imprecise Result or Readings	02/07/2023	02/07/2023
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A070909	Oversensing	02/07/2023	02/07/2023
									A0712	Pacing Problem	02/07/2023	02/07/2023
									A0908	Incorrect, Inadequate or Imprecise Result or Readings	02/07/2023	02/07/2023
							F08	Hospitalization or Prolonged Hospitalization	A070909	Oversensing	02/07/2023	02/07/2023
									A0712	Pacing Problem	02/07/2023	02/07/2023
									A0908	Incorrect, Inadequate or Imprecise Result or Readings	02/07/2023	02/07/2023
							F1001	Absence of Treatment	A070909	Oversensing	02/07/2023	02/07/2023
									A0712	Pacing Problem	02/07/2023	02/07/2023
									A0908	Incorrect, Inadequate or Imprecise Result or Readings	02/07/2023	02/07/2023
							F1905	Device Revision or Replacement	A070909	Oversensing	02/07/2023	02/07/2023
									A0712	Pacing Problem	02/07/2023	02/07/2023
									A0908	Incorrect, Inadequate or Imprecise Result or Readings	02/07/2023	02/07/2023
14333	2023	All other reportable incidents	37265	Dual-chamber implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E2104	Electric Shock	A072201	High Impedance		
							F17	Sedation	A072201	High Impedance		
							F22	Unexpected Diagnostic Intervention	A072201	High Impedance		
14334	2023	All other reportable incidents	47265	Dual-chamber implantable pacemaker, rate-responsive	CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	01/10/2023	01/10/2023
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	01/10/2023	01/10/2023
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	01/10/2023	01/10/2023
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	01/10/2023	01/10/2023
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	01/10/2023	01/10/2023
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	01/10/2023	01/10/2023
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	01/10/2023	01/10/2023
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	01/10/2023	01/10/2023
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	01/10/2023	01/10/2023
14335	2023	All other reportable incidents	47265	Dual-chamber implantable pacemaker, rate-responsive	CT1685	Cardiac pacemakers	E0112	Dizziness	A070910	Undersensing		
									A0712	Pacing Problem		
							E0601	Arrhythmia	A070910	Undersensing		
									A0712	Pacing Problem		
							E0717	Dyspnea	A070910	Undersensing		
									A0712	Pacing Problem		
							F23	Unexpected Medical Intervention	A070910	Undersensing		
									A0712	Pacing Problem		
					CT257	Implantable cardiac pacemakers	E0112	Dizziness	A070910	Undersensing		
									A0712	Pacing Problem		
							E0601	Arrhythmia	A070910	Undersensing		
									A0712	Pacing Problem		
							E0717	Dyspnea	A070910	Undersensing		
									A0712	Pacing Problem		

							F23	Unexpected Medical Intervention	A070910	Undersensing			
									A0712	Pacing Problem			
					CT2866	Implantable cardiac electrical stimulators	E0112	Dizziness	A070910	Undersensing			
									A0712	Pacing Problem			
							E0601	Arrhythmia	A070910	Undersensing			
									A0712	Pacing Problem			
							E0717	Dyspnea	A070910	Undersensing			
									A0712	Pacing Problem			
							F23	Unexpected Medical Intervention	A070910	Undersensing			
									A0712	Pacing Problem			
14336	2023	All other reportable incidents	47265	Dual-chamber implantable pacemaker, rate-responsive	CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem	06/10/2023	06/10/2023	
									A070504	Premature Discharge of Battery	06/10/2023	06/10/2023	
							F08	Hospitalization or Prolonged Hospitalization	A0705	Battery Problem	06/10/2023	06/10/2023	
									A070504	Premature Discharge of Battery	06/10/2023	06/10/2023	
							F1905	Device Revision or Replacement	A0705	Battery Problem	06/10/2023	06/10/2023	
									A070504	Premature Discharge of Battery	06/10/2023	06/10/2023	
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem	06/10/2023	06/10/2023	
									A070504	Premature Discharge of Battery	06/10/2023	06/10/2023	
							F08	Hospitalization or Prolonged Hospitalization	A0705	Battery Problem	06/10/2023	06/10/2023	
									A070504	Premature Discharge of Battery	06/10/2023	06/10/2023	
							F1905	Device Revision or Replacement	A0705	Battery Problem	06/10/2023	06/10/2023	
									A070504	Premature Discharge of Battery	06/10/2023	06/10/2023	
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem	06/10/2023	06/10/2023	
									A070504	Premature Discharge of Battery	06/10/2023	06/10/2023	
							F08	Hospitalization or Prolonged Hospitalization	A0705	Battery Problem	06/10/2023	06/10/2023	
									A070504	Premature Discharge of Battery	06/10/2023	06/10/2023	
							F1905	Device Revision or Replacement	A0705	Battery Problem	06/10/2023	06/10/2023	
									A070504	Premature Discharge of Battery	06/10/2023	06/10/2023	
14337	2023		47265	Dual-chamber implantable pacemaker, rate-responsive	CT141	Electrodes/Leads and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	20/09/2023	20/09/2023	
							F1901	Additional Surgery	A24	No Apparent Device Problem	20/09/2023	20/09/2023	
					CT145	Cardiovascular devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	20/09/2023	20/09/2023	
							F1901	Additional Surgery	A24	No Apparent Device Problem	20/09/2023	20/09/2023	
					CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	20/09/2023	20/09/2023	
							F1901	Additional Surgery	A24	No Apparent Device Problem	20/09/2023	20/09/2023	
					CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	20/09/2023	20/09/2023	
							F1901	Additional Surgery	A24	No Apparent Device Problem	20/09/2023	20/09/2023	
					CT233	Surgical	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	20/09/2023	20/09/2023	
							F1901	Additional Surgery	A24	No Apparent Device Problem	20/09/2023	20/09/2023	
					CT2406	Implantable	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	20/09/2023	20/09/2023	
							F1901	Additional Surgery	A24	No Apparent Device Problem	20/09/2023	20/09/2023	
					CT2407	Active implantable	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	20/09/2023	20/09/2023	
							F1901	Additional Surgery	A24	No Apparent Device Problem	20/09/2023	20/09/2023	
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	20/09/2023	20/09/2023	
							F1901	Additional Surgery	A24	No Apparent Device Problem	20/09/2023	20/09/2023	
					CT2727	Active implant devices/ancillaries	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	20/09/2023	20/09/2023	
							F1901	Additional Surgery	A24	No Apparent Device Problem	20/09/2023	20/09/2023	
					CT2750	Implantable cardiac electrical stimulators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	20/09/2023	20/09/2023	
							F1901	Additional Surgery	A24	No Apparent Device Problem	20/09/2023	20/09/2023	
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	20/09/2023	20/09/2023	
							F1901	Additional Surgery	A24	No Apparent Device Problem	20/09/2023	20/09/2023	
					CT2985	Active	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	20/09/2023	20/09/2023	
							F1901	Additional Surgery	A24	No Apparent Device Problem	20/09/2023	20/09/2023	
					CT3001	Implantable metal	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	20/09/2023	20/09/2023	
							F1901	Additional Surgery	A24	No Apparent Device Problem	20/09/2023	20/09/2023	
					CT3002	Implantable synthetic polymer	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	20/09/2023	20/09/2023	
							F1901	Additional Surgery	A24	No Apparent Device Problem	20/09/2023	20/09/2023	
					CT3006	Implantable inorganic material	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	20/09/2023	20/09/2023	
							F1901	Additional Surgery	A24	No Apparent Device Problem	20/09/2023	20/09/2023	
					CT314	Electrically powered	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	20/09/2023	20/09/2023	
							F1901	Additional Surgery	A24	No Apparent Device Problem	20/09/2023	20/09/2023	
					CT321	Long-term surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	20/09/2023	20/09/2023	
							F1901	Additional Surgery	A24	No Apparent Device Problem	20/09/2023	20/09/2023	
					CT449	Generators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	20/09/2023	20/09/2023	
							F1901	Additional Surgery	A24	No Apparent Device Problem	20/09/2023	20/09/2023	
					CT981	Single use	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	20/09/2023	20/09/2023	
							F1901	Additional Surgery	A24	No Apparent Device Problem	20/09/2023	20/09/2023	
					CT983	Surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	20/09/2023	20/09/2023	
							F1901	Additional Surgery	A24	No Apparent Device Problem	20/09/2023	20/09/2023	
14338	2023		47265	Dual-chamber implantable pacemaker, rate-responsive	CT141	Electrodes/Leads and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	31/10/2023	31/10/2023	
							F1901	Additional Surgery	A24	No Apparent Device Problem	31/10/2023	31/10/2023	
					CT145	Cardiovascular devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	31/10/2023	31/10/2023	
							F1901	Additional Surgery	A24	No Apparent Device Problem	31/10/2023	31/10/2023	
					CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	31/10/2023	31/10/2023	
							F1901	Additional Surgery	A24	No Apparent Device Problem	31/10/2023	31/10/2023	
					CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	31/10/2023	31/10/2023	
							F1901	Additional Surgery	A24	No Apparent Device Problem	31/10/2023	31/10/2023	
					CT233	Surgical	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	31/10/2023	31/10/2023	
							F1901	Additional Surgery	A24	No Apparent Device Problem	31/10/2023	31/10/2023	
					CT2406	Implantable	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	31/10/2023	31/10/2023	
							F1901	Additional Surgery	A24	No Apparent Device Problem	31/10/2023	31/10/2023	
					CT2407	Active implantable	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	31/10/2023	31/10/2023	
							F1901	Additional Surgery	A24	No Apparent Device Problem	31/10/2023	31/10/2023	
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	31/10/2023	31/10/2023	
							F1901	Additional Surgery	A24	No Apparent Device Problem	31/10/2023	31/10/2023	
					CT2727	Active implant devices/ancillaries	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	31/10/2023	31/10/2023	
							F1901	Additional Surgery	A24	No Apparent Device Problem	31/10/2023	31/10/2023	
					CT2750	Implantable cardiac electrical stimulators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	31/10/2023	31/10/2023	
							F1901	Additional Surgery	A24	No Apparent Device Problem	31/10/2023	31/10/2023	

					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	31/10/2023	31/10/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	31/10/2023	31/10/2023
					CT2985	Active	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	31/10/2023	31/10/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	31/10/2023	31/10/2023
					CT3001	Implantable metal	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	31/10/2023	31/10/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	31/10/2023	31/10/2023
					CT3002	Implantable synthetic polymer	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	31/10/2023	31/10/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	31/10/2023	31/10/2023
					CT3006	Implantable inorganic material	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	31/10/2023	31/10/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	31/10/2023	31/10/2023
					CT314	Electrically powered	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	31/10/2023	31/10/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	31/10/2023	31/10/2023
					CT321	Long-term surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	31/10/2023	31/10/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	31/10/2023	31/10/2023
					CT449	Generators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	31/10/2023	31/10/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	31/10/2023	31/10/2023
					CT981	Single use	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	31/10/2023	31/10/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	31/10/2023	31/10/2023
					CT983	Surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	31/10/2023	31/10/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	31/10/2023	31/10/2023
14339	2023		47265	Dual-chamber implantable pacemaker, rate-responsive	CT141	Electrodes/Leads and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	31/10/2023	31/10/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	31/10/2023	31/10/2023
					CT145	Cardiovascular devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	31/10/2023	31/10/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	31/10/2023	31/10/2023
					CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	31/10/2023	31/10/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	31/10/2023	31/10/2023
					CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	31/10/2023	31/10/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	31/10/2023	31/10/2023
					CT233	Surgical	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	31/10/2023	31/10/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	31/10/2023	31/10/2023
					CT2406	Implantable	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	31/10/2023	31/10/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	31/10/2023	31/10/2023
					CT2407	Active implantable	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	31/10/2023	31/10/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	31/10/2023	31/10/2023
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	31/10/2023	31/10/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	31/10/2023	31/10/2023
					CT2727	Active implant devices/ancillaries	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	31/10/2023	31/10/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	31/10/2023	31/10/2023
					CT2750	Implantable cardiac electrical stimulators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	31/10/2023	31/10/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	31/10/2023	31/10/2023
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	31/10/2023	31/10/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	31/10/2023	31/10/2023
					CT2985	Active	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	31/10/2023	31/10/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	31/10/2023	31/10/2023
					CT3001	Implantable metal	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	31/10/2023	31/10/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	31/10/2023	31/10/2023
					CT3002	Implantable synthetic polymer	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	31/10/2023	31/10/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	31/10/2023	31/10/2023
					CT3006	Implantable inorganic material	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	31/10/2023	31/10/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	31/10/2023	31/10/2023
					CT314	Electrically powered	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	31/10/2023	31/10/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	31/10/2023	31/10/2023
					CT321	Long-term surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	31/10/2023	31/10/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	31/10/2023	31/10/2023
					CT449	Generators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	31/10/2023	31/10/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	31/10/2023	31/10/2023
					CT981	Single use	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	31/10/2023	31/10/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	31/10/2023	31/10/2023
					CT983	Surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	31/10/2023	31/10/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	31/10/2023	31/10/2023
14340	2023	All other reportable incidents	47263	Cardiac resynchronization therapy implantable pacemaker	CT1685	Cardiac pacemakers	E060101	Asystole	A070909	Oversensing	01/08/2023	01/08/2023
									A0712	Pacing Problem	01/08/2023	01/08/2023
									A0908	Incorrect, Inadequate or Imprecise Result or Readings	01/08/2023	01/08/2023
									A090801	Signal Artifact/Noise	01/08/2023	01/08/2023
							F08	Hospitalization or Prolonged Hospitalization	A070909	Oversensing	01/08/2023	01/08/2023
									A0712	Pacing Problem	01/08/2023	01/08/2023
									A0908	Incorrect, Inadequate or Imprecise Result or Readings	01/08/2023	01/08/2023
									A090801	Signal Artifact/Noise	01/08/2023	01/08/2023
							F1001	Absence of Treatment	A070909	Oversensing	01/08/2023	01/08/2023
									A0712	Pacing Problem	01/08/2023	01/08/2023
									A0908	Incorrect, Inadequate or Imprecise Result or Readings	01/08/2023	01/08/2023
									A090801	Signal Artifact/Noise	01/08/2023	01/08/2023
							F1905	Device Revision or Replacement	A070909	Oversensing	01/08/2023	01/08/2023
									A0712	Pacing Problem	01/08/2023	01/08/2023
									A0908	Incorrect, Inadequate or Imprecise Result or Readings	01/08/2023	01/08/2023
									A090801	Signal Artifact/Noise	01/08/2023	01/08/2023
					CT257	Implantable cardiac pacemakers	E060101	Asystole	A070909	Oversensing	01/08/2023	01/08/2023
									A0712	Pacing Problem	01/08/2023	01/08/2023
									A0908	Incorrect, Inadequate or Imprecise Result or Readings	01/08/2023	01/08/2023
									A090801	Signal Artifact/Noise	01/08/2023	01/08/2023
							F08	Hospitalization or Prolonged Hospitalization	A070909	Oversensing	01/08/2023	01/08/2023
									A0712	Pacing Problem	01/08/2023	01/08/2023
									A0908	Incorrect, Inadequate or Imprecise Result or Readings	01/08/2023	01/08/2023

								A090801	Signal Artifact/Noise	01/08/2023	01/08/2023
						F1001	Absence of Treatment	A070909	Oversensing	01/08/2023	01/08/2023
								A0712	Pacing Problem	01/08/2023	01/08/2023
								A0908	Incorrect, Inadequate or Imprecise Result or Readings	01/08/2023	01/08/2023
								A090801	Signal Artifact/Noise	01/08/2023	01/08/2023
						F1905	Device Revision or Replacement	A070909	Oversensing	01/08/2023	01/08/2023
								A0712	Pacing Problem	01/08/2023	01/08/2023
								A0908	Incorrect, Inadequate or Imprecise Result or Readings	01/08/2023	01/08/2023
								A090801	Signal Artifact/Noise	01/08/2023	01/08/2023
				CT2866	Implantable cardiac electrical stimulators	E060101	Asystole	A070909	Oversensing	01/08/2023	01/08/2023
								A0712	Pacing Problem	01/08/2023	01/08/2023
								A0908	Incorrect, Inadequate or Imprecise Result or Readings	01/08/2023	01/08/2023
								A090801	Signal Artifact/Noise	01/08/2023	01/08/2023
						F08	Hospitalization or Prolonged Hospitalization	A070909	Oversensing	01/08/2023	01/08/2023
								A0712	Pacing Problem	01/08/2023	01/08/2023
								A0908	Incorrect, Inadequate or Imprecise Result or Readings	01/08/2023	01/08/2023
								A090801	Signal Artifact/Noise	01/08/2023	01/08/2023
						F1001	Absence of Treatment	A070909	Oversensing	01/08/2023	01/08/2023
								A0712	Pacing Problem	01/08/2023	01/08/2023
								A0908	Incorrect, Inadequate or Imprecise Result or Readings	01/08/2023	01/08/2023
								A090801	Signal Artifact/Noise	01/08/2023	01/08/2023
						F1905	Device Revision or Replacement	A070909	Oversensing	01/08/2023	01/08/2023
								A0712	Pacing Problem	01/08/2023	01/08/2023
								A0908	Incorrect, Inadequate or Imprecise Result or Readings	01/08/2023	01/08/2023
								A090801	Signal Artifact/Noise	01/08/2023	01/08/2023
14341	2023	All other reportable incidents	47265	Dual-chamber implantable pacemaker, rate-responsive	CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem	14/11/2023	14/11/2023
									A070504	Premature Discharge of Battery	14/11/2023	14/11/2023
							F08	Hospitalization or Prolonged Hospitalization	A0705	Battery Problem	14/11/2023	14/11/2023
									A070504	Premature Discharge of Battery	14/11/2023	14/11/2023
							F1905	Device Revision or Replacement	A0705	Battery Problem	14/11/2023	14/11/2023
									A070504	Premature Discharge of Battery	14/11/2023	14/11/2023
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem	14/11/2023	14/11/2023
									A070504	Premature Discharge of Battery	14/11/2023	14/11/2023
							F08	Hospitalization or Prolonged Hospitalization	A0705	Battery Problem	14/11/2023	14/11/2023
									A070504	Premature Discharge of Battery	14/11/2023	14/11/2023
							F1905	Device Revision or Replacement	A0705	Battery Problem	14/11/2023	14/11/2023
									A070504	Premature Discharge of Battery	14/11/2023	14/11/2023
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem	14/11/2023	14/11/2023
									A070504	Premature Discharge of Battery	14/11/2023	14/11/2023
							F08	Hospitalization or Prolonged Hospitalization	A0705	Battery Problem	14/11/2023	14/11/2023
									A070504	Premature Discharge of Battery	14/11/2023	14/11/2023
							F1905	Device Revision or Replacement	A0705	Battery Problem	14/11/2023	14/11/2023
									A070504	Premature Discharge of Battery	14/11/2023	14/11/2023
14342	2023	All other reportable incidents	47265	Dual-chamber implantable pacemaker, rate-responsive	CT141	Electrodes/Leads and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070909	Oversensing	27/10/2023	27/10/2023
									A0712	Pacing Problem	27/10/2023	27/10/2023
									A090801	Signal Artifact/Noise	27/10/2023	27/10/2023
							F1001	Absence of Treatment	A070909	Oversensing	27/10/2023	27/10/2023
									A0712	Pacing Problem	27/10/2023	27/10/2023
									A090801	Signal Artifact/Noise	27/10/2023	27/10/2023
							F1908	Prolonged Surgery	A070909	Oversensing	27/10/2023	27/10/2023
									A0712	Pacing Problem	27/10/2023	27/10/2023
									A090801	Signal Artifact/Noise	27/10/2023	27/10/2023
					CT145	Cardiovascular devices	E2403	No Clinical Signs, Symptoms or Conditions	A070909	Oversensing	27/10/2023	27/10/2023
									A0712	Pacing Problem	27/10/2023	27/10/2023
									A090801	Signal Artifact/Noise	27/10/2023	27/10/2023
							F1001	Absence of Treatment	A070909	Oversensing	27/10/2023	27/10/2023
									A0712	Pacing Problem	27/10/2023	27/10/2023
									A090801	Signal Artifact/Noise	27/10/2023	27/10/2023
							F1908	Prolonged Surgery	A070909	Oversensing	27/10/2023	27/10/2023
									A0712	Pacing Problem	27/10/2023	27/10/2023
									A090801	Signal Artifact/Noise	27/10/2023	27/10/2023
					CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070909	Oversensing	27/10/2023	27/10/2023
									A0712	Pacing Problem	27/10/2023	27/10/2023
									A090801	Signal Artifact/Noise	27/10/2023	27/10/2023
							F1001	Absence of Treatment	A070909	Oversensing	27/10/2023	27/10/2023
									A0712	Pacing Problem	27/10/2023	27/10/2023
									A090801	Signal Artifact/Noise	27/10/2023	27/10/2023
							F1908	Prolonged Surgery	A070909	Oversensing	27/10/2023	27/10/2023
									A0712	Pacing Problem	27/10/2023	27/10/2023
									A090801	Signal Artifact/Noise	27/10/2023	27/10/2023
					CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A070909	Oversensing	27/10/2023	27/10/2023
									A0712	Pacing Problem	27/10/2023	27/10/2023
									A090801	Signal Artifact/Noise	27/10/2023	27/10/2023
							F1001	Absence of Treatment	A070909	Oversensing	27/10/2023	27/10/2023
									A0712	Pacing Problem	27/10/2023	27/10/2023
									A090801	Signal Artifact/Noise	27/10/2023	27/10/2023
							F1908	Prolonged Surgery	A070909	Oversensing	27/10/2023	27/10/2023
									A0712	Pacing Problem	27/10/2023	27/10/2023
									A090801	Signal Artifact/Noise	27/10/2023	27/10/2023
					CT233	Surgical	E2403	No Clinical Signs, Symptoms or Conditions	A070909	Oversensing	27/10/2023	27/10/2023
									A0712	Pacing Problem	27/10/2023	27/10/2023
									A090801	Signal Artifact/Noise	27/10/2023	27/10/2023
							F1001	Absence of Treatment	A070909	Oversensing	27/10/2023	27/10/2023

									A0712	Pacing Problem	27/10/2023	27/10/2023
									A090801	Signal Artifact/Noise	27/10/2023	27/10/2023
							F1908	Prolonged Surgery	A070909	Oversensing	27/10/2023	27/10/2023
									A0712	Pacing Problem	27/10/2023	27/10/2023
									A090801	Signal Artifact/Noise	27/10/2023	27/10/2023
					CT2406	Implantable	E2403	No Clinical Signs, Symptoms or Conditions	A070909	Oversensing	27/10/2023	27/10/2023
									A0712	Pacing Problem	27/10/2023	27/10/2023
									A090801	Signal Artifact/Noise	27/10/2023	27/10/2023
							F1001	Absence of Treatment	A070909	Oversensing	27/10/2023	27/10/2023
									A0712	Pacing Problem	27/10/2023	27/10/2023
									A090801	Signal Artifact/Noise	27/10/2023	27/10/2023
							F1908	Prolonged Surgery	A070909	Oversensing	27/10/2023	27/10/2023
									A0712	Pacing Problem	27/10/2023	27/10/2023
									A090801	Signal Artifact/Noise	27/10/2023	27/10/2023
					CT2407	Active implantable	E2403	No Clinical Signs, Symptoms or Conditions	A070909	Oversensing	27/10/2023	27/10/2023
									A0712	Pacing Problem	27/10/2023	27/10/2023
									A090801	Signal Artifact/Noise	27/10/2023	27/10/2023
							F1001	Absence of Treatment	A070909	Oversensing	27/10/2023	27/10/2023
									A0712	Pacing Problem	27/10/2023	27/10/2023
									A090801	Signal Artifact/Noise	27/10/2023	27/10/2023
							F1908	Prolonged Surgery	A070909	Oversensing	27/10/2023	27/10/2023
									A0712	Pacing Problem	27/10/2023	27/10/2023
									A090801	Signal Artifact/Noise	27/10/2023	27/10/2023
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A070909	Oversensing	27/10/2023	27/10/2023
									A0712	Pacing Problem	27/10/2023	27/10/2023
									A090801	Signal Artifact/Noise	27/10/2023	27/10/2023
							F1001	Absence of Treatment	A070909	Oversensing	27/10/2023	27/10/2023
									A0712	Pacing Problem	27/10/2023	27/10/2023
									A090801	Signal Artifact/Noise	27/10/2023	27/10/2023
							F1908	Prolonged Surgery	A070909	Oversensing	27/10/2023	27/10/2023
									A0712	Pacing Problem	27/10/2023	27/10/2023
									A090801	Signal Artifact/Noise	27/10/2023	27/10/2023
					CT2727	Active implant devices/ancillaries	E2403	No Clinical Signs, Symptoms or Conditions	A070909	Oversensing	27/10/2023	27/10/2023
									A0712	Pacing Problem	27/10/2023	27/10/2023
									A090801	Signal Artifact/Noise	27/10/2023	27/10/2023
							F1001	Absence of Treatment	A070909	Oversensing	27/10/2023	27/10/2023
									A0712	Pacing Problem	27/10/2023	27/10/2023
									A090801	Signal Artifact/Noise	27/10/2023	27/10/2023
							F1908	Prolonged Surgery	A070909	Oversensing	27/10/2023	27/10/2023
									A0712	Pacing Problem	27/10/2023	27/10/2023
									A090801	Signal Artifact/Noise	27/10/2023	27/10/2023
					CT2750	Implantable cardiac electrical stimulators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070909	Oversensing	27/10/2023	27/10/2023
									A0712	Pacing Problem	27/10/2023	27/10/2023
									A090801	Signal Artifact/Noise	27/10/2023	27/10/2023
							F1001	Absence of Treatment	A070909	Oversensing	27/10/2023	27/10/2023
									A0712	Pacing Problem	27/10/2023	27/10/2023
									A090801	Signal Artifact/Noise	27/10/2023	27/10/2023
							F1908	Prolonged Surgery	A070909	Oversensing	27/10/2023	27/10/2023
									A0712	Pacing Problem	27/10/2023	27/10/2023
									A090801	Signal Artifact/Noise	27/10/2023	27/10/2023
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A070909	Oversensing	27/10/2023	27/10/2023
									A0712	Pacing Problem	27/10/2023	27/10/2023
									A090801	Signal Artifact/Noise	27/10/2023	27/10/2023
							F1001	Absence of Treatment	A070909	Oversensing	27/10/2023	27/10/2023
									A0712	Pacing Problem	27/10/2023	27/10/2023
									A090801	Signal Artifact/Noise	27/10/2023	27/10/2023
							F1908	Prolonged Surgery	A070909	Oversensing	27/10/2023	27/10/2023
									A0712	Pacing Problem	27/10/2023	27/10/2023
									A090801	Signal Artifact/Noise	27/10/2023	27/10/2023
					CT2985	Active	E2403	No Clinical Signs, Symptoms or Conditions	A070909	Oversensing	27/10/2023	27/10/2023
									A0712	Pacing Problem	27/10/2023	27/10/2023
									A090801	Signal Artifact/Noise	27/10/2023	27/10/2023
							F1001	Absence of Treatment	A070909	Oversensing	27/10/2023	27/10/2023
									A0712	Pacing Problem	27/10/2023	27/10/2023
									A090801	Signal Artifact/Noise	27/10/2023	27/10/2023
							F1908	Prolonged Surgery	A070909	Oversensing	27/10/2023	27/10/2023
									A0712	Pacing Problem	27/10/2023	27/10/2023
									A090801	Signal Artifact/Noise	27/10/2023	27/10/2023
					CT3001	Implantable metal	E2403	No Clinical Signs, Symptoms or Conditions	A070909	Oversensing	27/10/2023	27/10/2023
									A0712	Pacing Problem	27/10/2023	27/10/2023
									A090801	Signal Artifact/Noise	27/10/2023	27/10/2023
							F1001	Absence of Treatment	A070909	Oversensing	27/10/2023	27/10/2023
									A0712	Pacing Problem	27/10/2023	27/10/2023
									A090801	Signal Artifact/Noise	27/10/2023	27/10/2023
							F1908	Prolonged Surgery	A070909	Oversensing	27/10/2023	27/10/2023
									A0712	Pacing Problem	27/10/2023	27/10/2023
									A090801	Signal Artifact/Noise	27/10/2023	27/10/2023
					CT3002	Implantable synthetic polymer	E2403	No Clinical Signs, Symptoms or Conditions	A070909	Oversensing	27/10/2023	27/10/2023
									A0712	Pacing Problem	27/10/2023	27/10/2023
									A090801	Signal Artifact/Noise	27/10/2023	27/10/2023
							F1001	Absence of Treatment	A070909	Oversensing	27/10/2023	27/10/2023
									A0712	Pacing Problem	27/10/2023	27/10/2023
									A090801	Signal Artifact/Noise	27/10/2023	27/10/2023
							F1908	Prolonged Surgery	A070909	Oversensing	27/10/2023	27/10/2023
									A0712	Pacing Problem	27/10/2023	27/10/2023
									A090801	Signal Artifact/Noise	27/10/2023	27/10/2023
					CT3006	Implantable inorganic material	E2403	No Clinical Signs, Symptoms or Conditions	A070909	Oversensing	27/10/2023	27/10/2023
									A0712	Pacing Problem	27/10/2023	27/10/2023

									A090801	Signal Artifact/Noise	27/10/2023	27/10/2023
							F1001	Absence of Treatment	A070909	Oversensing	27/10/2023	27/10/2023
									A0712	Pacing Problem	27/10/2023	27/10/2023
									A090801	Signal Artifact/Noise	27/10/2023	27/10/2023
							F1908	Prolonged Surgery	A070909	Oversensing	27/10/2023	27/10/2023
									A0712	Pacing Problem	27/10/2023	27/10/2023
									A090801	Signal Artifact/Noise	27/10/2023	27/10/2023
					CT314	Electrically powered	E2403	No Clinical Signs, Symptoms or Conditions	A070909	Oversensing	27/10/2023	27/10/2023
									A0712	Pacing Problem	27/10/2023	27/10/2023
									A090801	Signal Artifact/Noise	27/10/2023	27/10/2023
							F1001	Absence of Treatment	A070909	Oversensing	27/10/2023	27/10/2023
									A0712	Pacing Problem	27/10/2023	27/10/2023
									A090801	Signal Artifact/Noise	27/10/2023	27/10/2023
							F1908	Prolonged Surgery	A070909	Oversensing	27/10/2023	27/10/2023
									A0712	Pacing Problem	27/10/2023	27/10/2023
									A090801	Signal Artifact/Noise	27/10/2023	27/10/2023
					CT321	Long-term surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A070909	Oversensing	27/10/2023	27/10/2023
									A0712	Pacing Problem	27/10/2023	27/10/2023
									A090801	Signal Artifact/Noise	27/10/2023	27/10/2023
							F1001	Absence of Treatment	A070909	Oversensing	27/10/2023	27/10/2023
									A0712	Pacing Problem	27/10/2023	27/10/2023
									A090801	Signal Artifact/Noise	27/10/2023	27/10/2023
							F1908	Prolonged Surgery	A070909	Oversensing	27/10/2023	27/10/2023
									A0712	Pacing Problem	27/10/2023	27/10/2023
									A090801	Signal Artifact/Noise	27/10/2023	27/10/2023
					CT449	Generators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070909	Oversensing	27/10/2023	27/10/2023
									A0712	Pacing Problem	27/10/2023	27/10/2023
									A090801	Signal Artifact/Noise	27/10/2023	27/10/2023
							F1001	Absence of Treatment	A070909	Oversensing	27/10/2023	27/10/2023
									A0712	Pacing Problem	27/10/2023	27/10/2023
									A090801	Signal Artifact/Noise	27/10/2023	27/10/2023
							F1908	Prolonged Surgery	A070909	Oversensing	27/10/2023	27/10/2023
									A0712	Pacing Problem	27/10/2023	27/10/2023
									A090801	Signal Artifact/Noise	27/10/2023	27/10/2023
					CT981	Single use	E2403	No Clinical Signs, Symptoms or Conditions	A070909	Oversensing	27/10/2023	27/10/2023
									A0712	Pacing Problem	27/10/2023	27/10/2023
									A090801	Signal Artifact/Noise	27/10/2023	27/10/2023
							F1001	Absence of Treatment	A070909	Oversensing	27/10/2023	27/10/2023
									A0712	Pacing Problem	27/10/2023	27/10/2023
									A090801	Signal Artifact/Noise	27/10/2023	27/10/2023
							F1908	Prolonged Surgery	A070909	Oversensing	27/10/2023	27/10/2023
									A0712	Pacing Problem	27/10/2023	27/10/2023
									A090801	Signal Artifact/Noise	27/10/2023	27/10/2023
					CT983	Surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A070909	Oversensing	27/10/2023	27/10/2023
									A0712	Pacing Problem	27/10/2023	27/10/2023
									A090801	Signal Artifact/Noise	27/10/2023	27/10/2023
							F1001	Absence of Treatment	A070909	Oversensing	27/10/2023	27/10/2023
									A0712	Pacing Problem	27/10/2023	27/10/2023
									A090801	Signal Artifact/Noise	27/10/2023	27/10/2023
							F1908	Prolonged Surgery	A070909	Oversensing	27/10/2023	27/10/2023
									A0712	Pacing Problem	27/10/2023	27/10/2023
									A090801	Signal Artifact/Noise	27/10/2023	27/10/2023
14343	2023	All other reportable incidents	17882	Physiologic-monitoring defibrillation system	CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070803	Failure to Power Up		
							F27	Problem Identified During Non-Clinical Procedure	A070803	Failure to Power Up		
14344	2023	All other reportable incidents	17882	Physiologic-monitoring defibrillation system	CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070803	Failure to Power Up		
							F27	Problem Identified During Non-Clinical Procedure	A070803	Failure to Power Up		
14345	2023		47265	Dual-chamber implantable pacemaker, rate-responsive	CT141	Electrodes/Leads and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	03/10/2023	03/10/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	03/10/2023	03/10/2023
					CT145	Cardiovascular devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	03/10/2023	03/10/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	03/10/2023	03/10/2023
					CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	03/10/2023	03/10/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	03/10/2023	03/10/2023
					CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	03/10/2023	03/10/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	03/10/2023	03/10/2023
					CT233	Surgical	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	03/10/2023	03/10/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	03/10/2023	03/10/2023
					CT2406	Implantable	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	03/10/2023	03/10/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	03/10/2023	03/10/2023
					CT2407	Active implantable	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	03/10/2023	03/10/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	03/10/2023	03/10/2023
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	03/10/2023	03/10/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	03/10/2023	03/10/2023
					CT2727	Active implant devices/ancillaries	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	03/10/2023	03/10/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	03/10/2023	03/10/2023
					CT2750	Implantable cardiac electrical stimulators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	03/10/2023	03/10/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	03/10/2023	03/10/2023
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	03/10/2023	03/10/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	03/10/2023	03/10/2023
					CT2985	Active	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	03/10/2023	03/10/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	03/10/2023	03/10/2023
					CT3001	Implantable metal	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	03/10/2023	03/10/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	03/10/2023	03/10/2023
					CT3002	Implantable synthetic polymer	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	03/10/2023	03/10/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	03/10/2023	03/10/2023
					CT3006	Implantable inorganic material	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	03/10/2023	03/10/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	03/10/2023	03/10/2023
					CT314	Electrically powered	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	03/10/2023	03/10/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	03/10/2023	03/10/2023

					CT321	Long-term surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	03/10/2023	03/10/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	03/10/2023	03/10/2023
					CT449	Generators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	03/10/2023	03/10/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	03/10/2023	03/10/2023
					CT981	Single use	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	03/10/2023	03/10/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	03/10/2023	03/10/2023
					CT983	Surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	03/10/2023	03/10/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	03/10/2023	03/10/2023
14346	2023	All other reportable incidents	37265	Dual-chamber implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E2104	Electric Shock	A071302	Inappropriate/Inadequate Shock/Stimulation	23/10/2023	23/10/2023
									A090801	Signal Artifact/Noise	23/10/2023	23/10/2023
									A1301	Failure to Read Input Signal	23/10/2023	23/10/2023
							F08	Hospitalization or Prolonged Hospitalization	A071302	Inappropriate/Inadequate Shock/Stimulation	23/10/2023	23/10/2023
									A090801	Signal Artifact/Noise	23/10/2023	23/10/2023
									A1301	Failure to Read Input Signal	23/10/2023	23/10/2023
							F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A071302	Inappropriate/Inadequate Shock/Stimulation	23/10/2023	23/10/2023
									A090801	Signal Artifact/Noise	23/10/2023	23/10/2023
									A1301	Failure to Read Input Signal	23/10/2023	23/10/2023
							F1905	Device Revision or Replacement	A071302	Inappropriate/Inadequate Shock/Stimulation	23/10/2023	23/10/2023
									A090801	Signal Artifact/Noise	23/10/2023	23/10/2023
									A1301	Failure to Read Input Signal	23/10/2023	23/10/2023
14347	2023	All other reportable incidents	47265	Dual-chamber implantable pacemaker, rate-responsive	CT1685	Cardiac pacemakers	E011903	Syncope/Fainting	A070504	Premature Discharge of Battery	20/10/2023	20/10/2023
									A070909	Oversensing	20/10/2023	20/10/2023
									A0712	Pacing Problem	20/10/2023	20/10/2023
									A0908	Incorrect, Inadequate or Imprecise Result or Readings	20/10/2023	20/10/2023
							E1610	Head Injury	A070504	Premature Discharge of Battery	20/10/2023	20/10/2023
									A070909	Oversensing	20/10/2023	20/10/2023
									A0712	Pacing Problem	20/10/2023	20/10/2023
									A0908	Incorrect, Inadequate or Imprecise Result or Readings	20/10/2023	20/10/2023
							E2007	Fall	A070504	Premature Discharge of Battery	20/10/2023	20/10/2023
									A070909	Oversensing	20/10/2023	20/10/2023
									A0712	Pacing Problem	20/10/2023	20/10/2023
									A0908	Incorrect, Inadequate or Imprecise Result or Readings	20/10/2023	20/10/2023
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	20/10/2023	20/10/2023
									A070909	Oversensing	20/10/2023	20/10/2023
									A0712	Pacing Problem	20/10/2023	20/10/2023
									A0908	Incorrect, Inadequate or Imprecise Result or Readings	20/10/2023	20/10/2023
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	20/10/2023	20/10/2023
									A070909	Oversensing	20/10/2023	20/10/2023
									A0712	Pacing Problem	20/10/2023	20/10/2023
									A0908	Incorrect, Inadequate or Imprecise Result or Readings	20/10/2023	20/10/2023
					CT257	Implantable cardiac pacemakers	E011903	Syncope/Fainting	A070504	Premature Discharge of Battery	20/10/2023	20/10/2023
									A070909	Oversensing	20/10/2023	20/10/2023
									A0712	Pacing Problem	20/10/2023	20/10/2023
									A0908	Incorrect, Inadequate or Imprecise Result or Readings	20/10/2023	20/10/2023
							E1610	Head Injury	A070504	Premature Discharge of Battery	20/10/2023	20/10/2023
									A070909	Oversensing	20/10/2023	20/10/2023
									A0712	Pacing Problem	20/10/2023	20/10/2023
									A0908	Incorrect, Inadequate or Imprecise Result or Readings	20/10/2023	20/10/2023
							E2007	Fall	A070504	Premature Discharge of Battery	20/10/2023	20/10/2023
									A070909	Oversensing	20/10/2023	20/10/2023
									A0712	Pacing Problem	20/10/2023	20/10/2023
									A0908	Incorrect, Inadequate or Imprecise Result or Readings	20/10/2023	20/10/2023
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	20/10/2023	20/10/2023
									A070909	Oversensing	20/10/2023	20/10/2023
									A0712	Pacing Problem	20/10/2023	20/10/2023
									A0908	Incorrect, Inadequate or Imprecise Result or Readings	20/10/2023	20/10/2023
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	20/10/2023	20/10/2023
									A070909	Oversensing	20/10/2023	20/10/2023
									A0712	Pacing Problem	20/10/2023	20/10/2023
									A0908	Incorrect, Inadequate or Imprecise Result or Readings	20/10/2023	20/10/2023
					CT2866	Implantable cardiac electrical stimulators	E011903	Syncope/Fainting	A070504	Premature Discharge of Battery	20/10/2023	20/10/2023
									A070909	Oversensing	20/10/2023	20/10/2023
									A0712	Pacing Problem	20/10/2023	20/10/2023
									A0908	Incorrect, Inadequate or Imprecise Result or Readings	20/10/2023	20/10/2023
							E1610	Head Injury	A070504	Premature Discharge of Battery	20/10/2023	20/10/2023
									A070909	Oversensing	20/10/2023	20/10/2023
									A0712	Pacing Problem	20/10/2023	20/10/2023
									A0908	Incorrect, Inadequate or Imprecise Result or Readings	20/10/2023	20/10/2023
							E2007	Fall	A070504	Premature Discharge of Battery	20/10/2023	20/10/2023
									A070909	Oversensing	20/10/2023	20/10/2023
									A0712	Pacing Problem	20/10/2023	20/10/2023
									A0908	Incorrect, Inadequate or Imprecise Result or Readings	20/10/2023	20/10/2023
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	20/10/2023	20/10/2023
									A070909	Oversensing	20/10/2023	20/10/2023
									A0712	Pacing Problem	20/10/2023	20/10/2023
									A0908	Incorrect, Inadequate or Imprecise Result or Readings	20/10/2023	20/10/2023
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	20/10/2023	20/10/2023

									A070909	Oversensing	20/10/2023	20/10/2023
									A0712	Pacing Problem	20/10/2023	20/10/2023
									A0908	Incorrect, Inadequate or Imprecise Result or Readings	20/10/2023	20/10/2023
14348	2023	All other reportable incidents	47265	Dual-chamber implantable pacemaker, rate-responsive	CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
14349	2023	All other reportable incidents	35852	Single-chamber implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
14350	2023	All other reportable incidents	47265	Dual-chamber implantable pacemaker, rate-responsive	CT141	Electrodes/Leads and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
					CT145	Cardiovascular devices	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
					CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
					CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
					CT233	Surgical	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
					CT2406	Implantable	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
					CT2407	Active implantable	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
					CT2727	Active implant devices/ancillaries	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
					CT2750	Implantable cardiac electrical stimulators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
					CT2985	Active	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
					CT3001	Implantable metal	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
					CT3002	Implantable synthetic polymer	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
					CT3006	Implantable inorganic material	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
					CT314	Electrically powered	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
					CT321	Long-term surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		

					CT449	Generators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem			
									A070504	Premature Discharge of Battery			
							F26	No Health Consequences or Impact	A0705	Battery Problem			
									A070504	Premature Discharge of Battery			
					CT981	Single use	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem			
									A070504	Premature Discharge of Battery			
							F26	No Health Consequences or Impact	A0705	Battery Problem			
									A070504	Premature Discharge of Battery			
					CT983	Surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem			
									A070504	Premature Discharge of Battery			
							F26	No Health Consequences or Impact	A0705	Battery Problem			
									A070504	Premature Discharge of Battery			
14351	2023	All other reportable incidents	47265	Dual-chamber implantable pacemaker, rate-responsive	CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A071102	Failure to Interrogate	24/10/2023	24/10/2023
							F1901	Additional Surgery	A071102	Failure to Interrogate	24/10/2023	24/10/2023
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A071102	Failure to Interrogate	24/10/2023	24/10/2023
							F1901	Additional Surgery	A071102	Failure to Interrogate	24/10/2023	24/10/2023
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A071102	Failure to Interrogate	24/10/2023	24/10/2023
							F1901	Additional Surgery	A071102	Failure to Interrogate	24/10/2023	24/10/2023
14352	2023	All other reportable incidents	47265	Dual-chamber implantable pacemaker, rate-responsive	CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	27/10/2023	27/10/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	27/10/2023	27/10/2023
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	27/10/2023	27/10/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	27/10/2023	27/10/2023
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	27/10/2023	27/10/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	27/10/2023	27/10/2023
14353	2023		47265	Dual-chamber implantable pacemaker, rate-responsive	CT141	Electrodes/Leads and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	31/10/2023	31/10/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	31/10/2023	31/10/2023
					CT145	Cardiovascular devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	31/10/2023	31/10/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	31/10/2023	31/10/2023
					CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	31/10/2023	31/10/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	31/10/2023	31/10/2023
					CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	31/10/2023	31/10/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	31/10/2023	31/10/2023
					CT233	Surgical	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	31/10/2023	31/10/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	31/10/2023	31/10/2023
					CT2406	Implantable	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	31/10/2023	31/10/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	31/10/2023	31/10/2023
					CT2407	Active implantable	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	31/10/2023	31/10/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	31/10/2023	31/10/2023
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	31/10/2023	31/10/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	31/10/2023	31/10/2023
					CT2727	Active implant devices/ancillaries	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	31/10/2023	31/10/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	31/10/2023	31/10/2023
					CT2750	Implantable cardiac electrical stimulators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	31/10/2023	31/10/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	31/10/2023	31/10/2023
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	31/10/2023	31/10/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	31/10/2023	31/10/2023
					CT2985	Active	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	31/10/2023	31/10/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	31/10/2023	31/10/2023
					CT3001	Implantable metal	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	31/10/2023	31/10/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	31/10/2023	31/10/2023
					CT3002	Implantable synthetic polymer	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	31/10/2023	31/10/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	31/10/2023	31/10/2023
					CT3006	Implantable inorganic material	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	31/10/2023	31/10/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	31/10/2023	31/10/2023
					CT314	Electrically powered	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	31/10/2023	31/10/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	31/10/2023	31/10/2023
					CT321	Long-term surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	31/10/2023	31/10/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	31/10/2023	31/10/2023
					CT449	Generators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	31/10/2023	31/10/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	31/10/2023	31/10/2023
					CT981	Single use	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	31/10/2023	31/10/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	31/10/2023	31/10/2023
					CT983	Surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	31/10/2023	31/10/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	31/10/2023	31/10/2023
14354	2023	All other reportable incidents	17882	Physiologic-monitoring defibrillation system	CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070603	Failure to Charge		
									A071204	No Pacing		
							F27	Problem Identified During Non-Clinical Procedure	A070603	Failure to Charge		
									A071204	No Pacing		
14355	2023	All other reportable incidents	47265	Dual-chamber implantable pacemaker, rate-responsive	CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	21/11/2023	21/11/2023
									A160601	Premature Elective Replacement Indicator	21/11/2023	21/11/2023
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	21/11/2023	21/11/2023
									A160601	Premature Elective Replacement Indicator	21/11/2023	21/11/2023
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	21/11/2023	21/11/2023
									A160601	Premature Elective Replacement Indicator	21/11/2023	21/11/2023
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	21/11/2023	21/11/2023
									A160601	Premature Elective Replacement Indicator	21/11/2023	21/11/2023
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	21/11/2023	21/11/2023
									A160601	Premature Elective Replacement Indicator	21/11/2023	21/11/2023
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	21/11/2023	21/11/2023
									A160601	Premature Elective Replacement Indicator	21/11/2023	21/11/2023
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	21/11/2023	21/11/2023
									A160601	Premature Elective Replacement Indicator	21/11/2023	21/11/2023
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	21/11/2023	21/11/2023
									A160601	Premature Elective Replacement Indicator	21/11/2023	21/11/2023
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	21/11/2023	21/11/2023
									A160601	Premature Elective Replacement Indicator	21/11/2023	21/11/2023
14356	2023	All other reportable incidents	47265	Dual-chamber implantable pacemaker, rate-responsive	CT141	Electrodes/Leads and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		

							F26	No Health Consequences or Impact	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
					CT145	Cardiovascular devices	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
					CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
					CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
					CT233	Surgical	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
					CT2406	Implantable	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
					CT2407	Active implantable	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
					CT2727	Active implant devices/ancillaries	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
					CT2750	Implantable cardiac electrical stimulators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
					CT2985	Active	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
					CT3001	Implantable metal	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
					CT3002	Implantable synthetic polymer	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
					CT3006	Implantable inorganic material	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
					CT314	Electrically powered	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
					CT321	Long-term surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
					CT449	Generators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
					CT981	Single use	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
					CT983	Surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
14357	2023	All other reportable incidents	17882	Physiologic-monitoring defibrillation system	CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070803	Failure to Power Up		
							F27	Problem Identified During Non-Clinical Procedure	A070803	Failure to Power Up		
14358	2023	All other reportable incidents	47265	Dual-chamber implantable pacemaker, rate-responsive	CT141	Electrodes/Leads and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem	12/05/2025	12/05/2025
									A070504	Premature Discharge of Battery	12/05/2025	12/05/2025
							F26	No Health Consequences or Impact	A0705	Battery Problem	12/05/2025	12/05/2025
									A070504	Premature Discharge of Battery	12/05/2025	12/05/2025
					CT145	Cardiovascular devices	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem	12/05/2025	12/05/2025
									A070504	Premature Discharge of Battery	12/05/2025	12/05/2025
							F26	No Health Consequences or Impact	A0705	Battery Problem	12/05/2025	12/05/2025
									A070504	Premature Discharge of Battery	12/05/2025	12/05/2025

					CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem	12/05/2025	12/05/2025
									A070504	Premature Discharge of Battery	12/05/2025	12/05/2025
							F26	No Health Consequences or Impact	A0705	Battery Problem	12/05/2025	12/05/2025
									A070504	Premature Discharge of Battery	12/05/2025	12/05/2025
					CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem	12/05/2025	12/05/2025
									A070504	Premature Discharge of Battery	12/05/2025	12/05/2025
							F26	No Health Consequences or Impact	A0705	Battery Problem	12/05/2025	12/05/2025
									A070504	Premature Discharge of Battery	12/05/2025	12/05/2025
					CT233	Surgical	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem	12/05/2025	12/05/2025
									A070504	Premature Discharge of Battery	12/05/2025	12/05/2025
							F26	No Health Consequences or Impact	A0705	Battery Problem	12/05/2025	12/05/2025
									A070504	Premature Discharge of Battery	12/05/2025	12/05/2025
					CT2406	Implantable	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem	12/05/2025	12/05/2025
									A070504	Premature Discharge of Battery	12/05/2025	12/05/2025
							F26	No Health Consequences or Impact	A0705	Battery Problem	12/05/2025	12/05/2025
									A070504	Premature Discharge of Battery	12/05/2025	12/05/2025
					CT2407	Active implantable	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem	12/05/2025	12/05/2025
									A070504	Premature Discharge of Battery	12/05/2025	12/05/2025
							F26	No Health Consequences or Impact	A0705	Battery Problem	12/05/2025	12/05/2025
									A070504	Premature Discharge of Battery	12/05/2025	12/05/2025
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem	12/05/2025	12/05/2025
									A070504	Premature Discharge of Battery	12/05/2025	12/05/2025
							F26	No Health Consequences or Impact	A0705	Battery Problem	12/05/2025	12/05/2025
									A070504	Premature Discharge of Battery	12/05/2025	12/05/2025
					CT2727	Active implant devices/ancillaries	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem	12/05/2025	12/05/2025
									A070504	Premature Discharge of Battery	12/05/2025	12/05/2025
							F26	No Health Consequences or Impact	A0705	Battery Problem	12/05/2025	12/05/2025
									A070504	Premature Discharge of Battery	12/05/2025	12/05/2025
					CT2750	Implantable cardiac electrical stimulators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem	12/05/2025	12/05/2025
									A070504	Premature Discharge of Battery	12/05/2025	12/05/2025
							F26	No Health Consequences or Impact	A0705	Battery Problem	12/05/2025	12/05/2025
									A070504	Premature Discharge of Battery	12/05/2025	12/05/2025
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem	12/05/2025	12/05/2025
									A070504	Premature Discharge of Battery	12/05/2025	12/05/2025
							F26	No Health Consequences or Impact	A0705	Battery Problem	12/05/2025	12/05/2025
									A070504	Premature Discharge of Battery	12/05/2025	12/05/2025
					CT2985	Active	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem	12/05/2025	12/05/2025
									A070504	Premature Discharge of Battery	12/05/2025	12/05/2025
							F26	No Health Consequences or Impact	A0705	Battery Problem	12/05/2025	12/05/2025
									A070504	Premature Discharge of Battery	12/05/2025	12/05/2025
					CT3001	Implantable metal	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem	12/05/2025	12/05/2025
									A070504	Premature Discharge of Battery	12/05/2025	12/05/2025
							F26	No Health Consequences or Impact	A0705	Battery Problem	12/05/2025	12/05/2025
									A070504	Premature Discharge of Battery	12/05/2025	12/05/2025
					CT3002	Implantable synthetic polymer	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem	12/05/2025	12/05/2025
									A070504	Premature Discharge of Battery	12/05/2025	12/05/2025
							F26	No Health Consequences or Impact	A0705	Battery Problem	12/05/2025	12/05/2025
									A070504	Premature Discharge of Battery	12/05/2025	12/05/2025
					CT3006	Implantable inorganic material	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem	12/05/2025	12/05/2025
									A070504	Premature Discharge of Battery	12/05/2025	12/05/2025
							F26	No Health Consequences or Impact	A0705	Battery Problem	12/05/2025	12/05/2025
									A070504	Premature Discharge of Battery	12/05/2025	12/05/2025
					CT314	Electrically powered	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem	12/05/2025	12/05/2025
									A070504	Premature Discharge of Battery	12/05/2025	12/05/2025
							F26	No Health Consequences or Impact	A0705	Battery Problem	12/05/2025	12/05/2025
									A070504	Premature Discharge of Battery	12/05/2025	12/05/2025
					CT321	Long-term surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem	12/05/2025	12/05/2025
									A070504	Premature Discharge of Battery	12/05/2025	12/05/2025
							F26	No Health Consequences or Impact	A0705	Battery Problem	12/05/2025	12/05/2025
									A070504	Premature Discharge of Battery	12/05/2025	12/05/2025
					CT449	Generators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem	12/05/2025	12/05/2025
									A070504	Premature Discharge of Battery	12/05/2025	12/05/2025
							F26	No Health Consequences or Impact	A0705	Battery Problem	12/05/2025	12/05/2025
									A070504	Premature Discharge of Battery	12/05/2025	12/05/2025
					CT981	Single use	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem	12/05/2025	12/05/2025
									A070504	Premature Discharge of Battery	12/05/2025	12/05/2025
							F26	No Health Consequences or Impact	A0705	Battery Problem	12/05/2025	12/05/2025
									A070504	Premature Discharge of Battery	12/05/2025	12/05/2025
					CT983	Surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem	12/05/2025	12/05/2025
									A070504	Premature Discharge of Battery	12/05/2025	12/05/2025
							F26	No Health Consequences or Impact	A0705	Battery Problem	12/05/2025	12/05/2025
									A070504	Premature Discharge of Battery	12/05/2025	12/05/2025
14359	2023	All other reportable incidents	47265	Dual-chamber implantable pacemaker, rate-responsive	CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A0908	Incorrect, Inadequate or Imprecise Result or Readings		
							F08	Hospitalization or Prolonged Hospitalization	A0908	Incorrect, Inadequate or Imprecise Result or Readings		
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A0908	Incorrect, Inadequate or Imprecise Result or Readings		
							F08	Hospitalization or Prolonged Hospitalization	A0908	Incorrect, Inadequate or Imprecise Result or Readings		
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A0908	Incorrect, Inadequate or Imprecise Result or Readings		
							F08	Hospitalization or Prolonged Hospitalization	A0908	Incorrect, Inadequate or Imprecise Result or Readings		
14360	2023	All other reportable incidents	35852	Single-chamber implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
									A0908	Incorrect, Inadequate or Imprecise Result or Readings		
							F24	Insufficient Health Impact Information	A070504	Premature Discharge of Battery		
									A0908	Incorrect, Inadequate or Imprecise Result or Readings		

14361	2023	All other reportable incidents	35852	Single-chamber implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	23/11/2023	23/11/2023
									A0908	Incorrect, Inadequate or Imprecise Result or Readings	23/11/2023	23/11/2023
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	23/11/2023	23/11/2023
									A0908	Incorrect, Inadequate or Imprecise Result or Readings	23/11/2023	23/11/2023
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	23/11/2023	23/11/2023
									A0908	Incorrect, Inadequate or Imprecise Result or Readings	23/11/2023	23/11/2023
14362	2023		47265	Dual-chamber implantable pacemaker, rate-responsive	CT141	Electrodes/Leads and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	31/10/2023	31/10/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	31/10/2023	31/10/2023
					CT145	Cardiovascular devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	31/10/2023	31/10/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	31/10/2023	31/10/2023
					CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	31/10/2023	31/10/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	31/10/2023	31/10/2023
					CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	31/10/2023	31/10/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	31/10/2023	31/10/2023
					CT233	Surgical	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	31/10/2023	31/10/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	31/10/2023	31/10/2023
					CT2406	Implantable	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	31/10/2023	31/10/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	31/10/2023	31/10/2023
					CT2407	Active implantable	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	31/10/2023	31/10/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	31/10/2023	31/10/2023
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	31/10/2023	31/10/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	31/10/2023	31/10/2023
					CT2727	Active implant devices/ancillaries	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	31/10/2023	31/10/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	31/10/2023	31/10/2023
					CT2750	Implantable cardiac electrical stimulators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	31/10/2023	31/10/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	31/10/2023	31/10/2023
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	31/10/2023	31/10/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	31/10/2023	31/10/2023
					CT2985	Active	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	31/10/2023	31/10/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	31/10/2023	31/10/2023
					CT3001	Implantable metal	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	31/10/2023	31/10/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	31/10/2023	31/10/2023
					CT3002	Implantable synthetic polymer	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	31/10/2023	31/10/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	31/10/2023	31/10/2023
					CT3006	Implantable inorganic material	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	31/10/2023	31/10/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	31/10/2023	31/10/2023
					CT314	Electrically powered	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	31/10/2023	31/10/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	31/10/2023	31/10/2023
					CT321	Long-term surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	31/10/2023	31/10/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	31/10/2023	31/10/2023
					CT449	Generators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	31/10/2023	31/10/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	31/10/2023	31/10/2023
					CT981	Single use	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	31/10/2023	31/10/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	31/10/2023	31/10/2023
					CT983	Surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	31/10/2023	31/10/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	31/10/2023	31/10/2023
14363	2023	All other reportable incidents	47265	Dual-chamber implantable pacemaker, rate-responsive	CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	20/03/2024	20/03/2024
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	20/03/2024	20/03/2024
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	20/03/2024	20/03/2024
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	20/03/2024	20/03/2024
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	20/03/2024	20/03/2024
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	20/03/2024	20/03/2024
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	20/03/2024	20/03/2024
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	20/03/2024	20/03/2024
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	20/03/2024	20/03/2024
14364	2023	All other reportable incidents	17882	Physiologic-monitoring defibrillation system	CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A1501	Activation Problem		
							F27	Problem Identified During Non-Clinical Procedure	A1501	Activation Problem		
14365	2023	All other reportable incidents	17882	Physiologic-monitoring defibrillation system	CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070803	Failure to Power Up		
							F27	Problem Identified During Non-Clinical Procedure	A070803	Failure to Power Up		
14366	2023	All other reportable incidents	47265	Dual-chamber implantable pacemaker, rate-responsive	CT141	Electrodes/Leads and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	26/05/2023	26/05/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	26/05/2023	26/05/2023
					CT145	Cardiovascular devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	26/05/2023	26/05/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	26/05/2023	26/05/2023
					CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	26/05/2023	26/05/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	26/05/2023	26/05/2023
					CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	26/05/2023	26/05/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	26/05/2023	26/05/2023
					CT233	Surgical	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	26/05/2023	26/05/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	26/05/2023	26/05/2023
					CT2406	Implantable	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	26/05/2023	26/05/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	26/05/2023	26/05/2023
					CT2407	Active implantable	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	26/05/2023	26/05/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	26/05/2023	26/05/2023
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	26/05/2023	26/05/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	26/05/2023	26/05/2023
					CT2727	Active implant devices/ancillaries	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	26/05/2023	26/05/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	26/05/2023	26/05/2023
					CT2750	Implantable cardiac electrical stimulators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	26/05/2023	26/05/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	26/05/2023	26/05/2023
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	26/05/2023	26/05/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	26/05/2023	26/05/2023
					CT2985	Active	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	26/05/2023	26/05/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	26/05/2023	26/05/2023
					CT3001	Implantable metal	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	26/05/2023	26/05/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	26/05/2023	26/05/2023

					CT3002	Implantable synthetic polymer	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	26/05/2023	26/05/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	26/05/2023	26/05/2023
					CT3006	Implantable inorganic material	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	26/05/2023	26/05/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	26/05/2023	26/05/2023
					CT314	Electrically powered	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	26/05/2023	26/05/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	26/05/2023	26/05/2023
					CT321	Long-term surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	26/05/2023	26/05/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	26/05/2023	26/05/2023
					CT449	Generators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	26/05/2023	26/05/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	26/05/2023	26/05/2023
					CT981	Single use	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	26/05/2023	26/05/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	26/05/2023	26/05/2023
					CT983	Surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	26/05/2023	26/05/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	26/05/2023	26/05/2023
14367	2023	All other reportable incidents	35852	Single-chamber implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
									A0908	Incorrect, Inadequate or Imprecise Result or Readings		
							F24	Insufficient Health Impact Information	A070504	Premature Discharge of Battery		
									A0908	Incorrect, Inadequate or Imprecise Result or Readings		
14368	2023	All other reportable incidents	37265	Dual-chamber implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	01/12/2023	31/12/2023
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	01/12/2023	31/12/2023
							F12	Serious Injury/ Illness/ Impairment	A070504	Premature Discharge of Battery	01/12/2023	31/12/2023
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	01/12/2023	31/12/2023
14369	2023	All other reportable incidents	47265	Dual-chamber implantable pacemaker, rate-responsive	CT1685	Cardiac pacemakers	E0601	Arrhythmia	A071102	Failure to Interrogate	24/10/2023	24/10/2023
									A071204	No Pacing	24/10/2023	24/10/2023
							E0602	Cardiac Arrest	A071102	Failure to Interrogate	24/10/2023	24/10/2023
									A071204	No Pacing	24/10/2023	24/10/2023
							F1901	Additional Surgery	A071102	Failure to Interrogate	24/10/2023	24/10/2023
									A071204	No Pacing	24/10/2023	24/10/2023
					CT257	Implantable cardiac pacemakers	E0601	Arrhythmia	A071102	Failure to Interrogate	24/10/2023	24/10/2023
									A071204	No Pacing	24/10/2023	24/10/2023
							E0602	Cardiac Arrest	A071102	Failure to Interrogate	24/10/2023	24/10/2023
									A071204	No Pacing	24/10/2023	24/10/2023
							F1901	Additional Surgery	A071102	Failure to Interrogate	24/10/2023	24/10/2023
									A071204	No Pacing	24/10/2023	24/10/2023
					CT2866	Implantable cardiac electrical stimulators	E0601	Arrhythmia	A071102	Failure to Interrogate	24/10/2023	24/10/2023
									A071204	No Pacing	24/10/2023	24/10/2023
							E0602	Cardiac Arrest	A071102	Failure to Interrogate	24/10/2023	24/10/2023
									A071204	No Pacing	24/10/2023	24/10/2023
							F1901	Additional Surgery	A071102	Failure to Interrogate	24/10/2023	24/10/2023
									A071204	No Pacing	24/10/2023	24/10/2023
14370	2023	All other reportable incidents	35223	Endocardial/interventricular septal pacing lead	CT1684	Cardiac pacemakers and associated devices	E011903	Syncope/Fainting	A040101	Fracture	16/08/2023	16/08/2023
									A070909	Oversensing	16/08/2023	16/08/2023
									A0712	Pacing Problem	16/08/2023	16/08/2023
									A1205	Loose or Intermittent Connection	16/08/2023	16/08/2023
							E060101	Asystole	A040101	Fracture	16/08/2023	16/08/2023
									A070909	Oversensing	16/08/2023	16/08/2023
									A0712	Pacing Problem	16/08/2023	16/08/2023
									A1205	Loose or Intermittent Connection	16/08/2023	16/08/2023
							F08	Hospitalization or Prolonged Hospitalization	A040101	Fracture	16/08/2023	16/08/2023
									A070909	Oversensing	16/08/2023	16/08/2023
									A0712	Pacing Problem	16/08/2023	16/08/2023
									A1205	Loose or Intermittent Connection	16/08/2023	16/08/2023
							F1203	Life Threatening Illness or Injury	A040101	Fracture	16/08/2023	16/08/2023
									A070909	Oversensing	16/08/2023	16/08/2023
									A0712	Pacing Problem	16/08/2023	16/08/2023
									A1205	Loose or Intermittent Connection	16/08/2023	16/08/2023
							F1905	Device Revision or Replacement	A040101	Fracture	16/08/2023	16/08/2023
									A070909	Oversensing	16/08/2023	16/08/2023
									A0712	Pacing Problem	16/08/2023	16/08/2023
									A1205	Loose or Intermittent Connection	16/08/2023	16/08/2023
14371	2023	All other reportable incidents	17882	Physiologic-monitoring defibrillation system	CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070803	Failure to Power Up		
							F27	Problem Identified During Non-Clinical Procedure	A070803	Failure to Power Up		
14372	2023	All other reportable incidents	17882	Physiologic-monitoring defibrillation system	CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070803	Failure to Power Up		
							F27	Problem Identified During Non-Clinical Procedure	A070803	Failure to Power Up		
14373	2023	All other reportable incidents	35852	Single-chamber implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A070909	Oversensing		
									A090801	Signal Artifact/Noise		
							F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A070909	Oversensing		
									A090801	Signal Artifact/Noise		
							F2203	Imaging Required	A070909	Oversensing		
									A090801	Signal Artifact/Noise		
							F23	Unexpected Medical Intervention	A070909	Oversensing		
									A090801	Signal Artifact/Noise		
14374	2023	All other reportable incidents	47265	Dual-chamber implantable pacemaker, rate-responsive	CT141	Electrodes/Leads and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
					CT145	Cardiovascular devices	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
					CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
					CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		

							F26	No Health Consequences or Impact	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
					CT233	Surgical	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
					CT2406	Implantable	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
					CT2407	Active implantable	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
					CT2727	Active implant devices/ancillaries	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
					CT2750	Implantable cardiac electrical stimulators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
					CT2985	Active	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
					CT3001	Implantable metal	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
					CT3002	Implantable synthetic polymer	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
					CT3006	Implantable inorganic material	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
					CT314	Electrically powered	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
					CT321	Long-term surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
					CT449	Generators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
					CT981	Single use	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
					CT983	Surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
14375	2023	All other reportable incidents	35223	Endocardial/interventricular septal pacing lead	CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A0401	Break		
									A070909	Oversensing		
									A090801	Signal Artifact/Noise		
							F26	No Health Consequences or Impact	A0401	Break		
									A070909	Oversensing		
									A090801	Signal Artifact/Noise		
14376	2023	All other reportable incidents	60789	Intracardiac pacemaker	CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A051201	Device Dislodged or Dislocated	03/11/2023	03/11/2023
							F26	No Health Consequences or Impact	A051201	Device Dislodged or Dislocated	03/11/2023	03/11/2023
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A051201	Device Dislodged or Dislocated	03/11/2023	03/11/2023
							F26	No Health Consequences or Impact	A051201	Device Dislodged or Dislocated	03/11/2023	03/11/2023
					CT2818	Catheterization/Cannulation/Fluid path devices	E2403	No Clinical Signs, Symptoms or Conditions	A051201	Device Dislodged or Dislocated	03/11/2023	03/11/2023
							F26	No Health Consequences or Impact	A051201	Device Dislodged or Dislocated	03/11/2023	03/11/2023
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A051201	Device Dislodged or Dislocated	03/11/2023	03/11/2023
							F26	No Health Consequences or Impact	A051201	Device Dislodged or Dislocated	03/11/2023	03/11/2023
14377	2023	All other reportable incidents	47265	Dual-chamber implantable pacemaker, rate-responsive	CT141	Electrodes/Leads and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT145	Cardiovascular devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		

					CT233	Surgical	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT2406	Implantable	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT2407	Active implantable	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT2727	Active implant devices/ancillaries	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT2750	Implantable cardiac electrical stimulators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT2985	Active	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT3001	Implantable metal	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT3002	Implantable synthetic polymer	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT3006	Implantable inorganic material	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT314	Electrically powered	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT321	Long-term surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT449	Generators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT981	Single use	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT983	Surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
14378	2023	All other reportable incidents	36241	Pacemaker/defibrillator lead	CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070909	Oversensing		
							F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A070909	Oversensing		
							F2203	Imaging Required	A070909	Oversensing		
							F23	Unexpected Medical Intervention	A070909	Oversensing		
14379	2023	All other reportable incidents	47265	Dual-chamber implantable pacemaker, rate-responsive	CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem	27/10/2023	27/10/2023
							F08	Hospitalization or Prolonged Hospitalization	A0705	Battery Problem	27/10/2023	27/10/2023
							F1905	Device Revision or Replacement	A0705	Battery Problem	27/10/2023	27/10/2023
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem	27/10/2023	27/10/2023
							F08	Hospitalization or Prolonged Hospitalization	A0705	Battery Problem	27/10/2023	27/10/2023
							F1905	Device Revision or Replacement	A0705	Battery Problem	27/10/2023	27/10/2023
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem	27/10/2023	27/10/2023
							F08	Hospitalization or Prolonged Hospitalization	A0705	Battery Problem	27/10/2023	27/10/2023
							F1905	Device Revision or Replacement	A0705	Battery Problem	27/10/2023	27/10/2023
14380	2023	All other reportable incidents	36241	Pacemaker/defibrillator lead	CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070909	Oversensing		
									A090801	Signal Artifact/Noise		
							F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A070909	Oversensing		
									A090801	Signal Artifact/Noise		
							F2203	Imaging Required	A070909	Oversensing		
									A090801	Signal Artifact/Noise		
							F23	Unexpected Medical Intervention	A070909	Oversensing		
									A090801	Signal Artifact/Noise		
14381	2023	All other reportable incidents	17882	Physiologic-monitoring defibrillation system	CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A150105	Key or Button Unresponsive/not Working		
							F27	Problem Identified During Non-Clinical Procedure	A150105	Key or Button Unresponsive/not Working		
14382	2023	All other reportable incidents	17882	Physiologic-monitoring defibrillation system	CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A07	Electrical /Electronic Property Problem		
							F27	Problem Identified During Non-Clinical Procedure	A07	Electrical /Electronic Property Problem		
14383	2023	All other reportable incidents	37265	Dual-chamber implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E0601	Arrhythmia	A071201	Failure to Convert Rhythm		
							F12	Serious Injury/ Illness/ Impairment	A071201	Failure to Convert Rhythm		
14384	2023	All other reportable incidents	35852	Single-chamber implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	15/09/2023	15/09/2023
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	15/09/2023	15/09/2023
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	15/09/2023	15/09/2023
14385	2023	All other reportable incidents	47265	Dual-chamber implantable pacemaker, rate-responsive	CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	01/11/2023	01/11/2023
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	01/11/2023	01/11/2023
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	01/11/2023	01/11/2023
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	01/11/2023	01/11/2023
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	01/11/2023	01/11/2023
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	01/11/2023	01/11/2023
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	01/11/2023	01/11/2023
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	01/11/2023	01/11/2023
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	01/11/2023	01/11/2023
14386	2023	All other reportable incidents	35852	Single-chamber implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	11/03/2024	11/03/2024
									A0908	Incorrect, Inadequate or Imprecise Result or Readings	11/03/2024	11/03/2024
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	11/03/2024	11/03/2024
									A0908	Incorrect, Inadequate or Imprecise Result or Readings	11/03/2024	11/03/2024
							F1903	Device Explantation	A070504	Premature Discharge of Battery	11/03/2024	11/03/2024
									A0908	Incorrect, Inadequate or Imprecise Result or Readings	11/03/2024	11/03/2024
14387	2023	All other reportable incidents	17882	Physiologic-monitoring defibrillation system	CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070803	Failure to Power Up		
							F27	Problem Identified During Non-Clinical Procedure	A070803	Failure to Power Up		
14388	2023	All other reportable incidents	47265	Dual-chamber implantable pacemaker, rate-responsive	CT141	Electrodes/Leads and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
					CT145	Cardiovascular devices	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		

							F26	No Health Consequences or Impact	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
					CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
					CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
					CT233	Surgical	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
					CT2406	Implantable	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
					CT2407	Active implantable	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
					CT2727	Active implant devices/ancillaries	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
					CT2750	Implantable cardiac electrical stimulators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
					CT2985	Active	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
					CT3001	Implantable metal	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
					CT3002	Implantable synthetic polymer	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
					CT3006	Implantable inorganic material	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
					CT314	Electrically powered	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
					CT321	Long-term surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
					CT449	Generators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
					CT981	Single use	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
					CT983	Surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
14389	2023	All other reportable incidents	35852	Single-chamber implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E011903	Syncope/Fainting	A070909	Oversensing		
									A071302	Inappropriate/Inadequate Shock/Stimulation		
							E2104	Electric Shock	A070909	Oversensing		
									A071302	Inappropriate/Inadequate Shock/Stimulation		
							E2311	Discomfort	A070909	Oversensing		
									A071302	Inappropriate/Inadequate Shock/Stimulation		
							E2330	Pain	A070909	Oversensing		
									A071302	Inappropriate/Inadequate Shock/Stimulation		
							F08	Hospitalization or Prolonged Hospitalization	A070909	Oversensing		
									A071302	Inappropriate/Inadequate Shock/Stimulation		
							F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A070909	Oversensing		
									A071302	Inappropriate/Inadequate Shock/Stimulation		
14390	2023	All other reportable incidents	35852	Single-chamber implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A0709	Device Sensing Problem	23/05/2023	23/05/2023
									A070909	Oversensing	23/05/2023	23/05/2023

							F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A0709	Device Sensing Problem	23/05/2023	23/05/2023
									A070909	Oversensing	23/05/2023	23/05/2023
							F2203	Imaging Required	A0709	Device Sensing Problem	23/05/2023	23/05/2023
									A070909	Oversensing	23/05/2023	23/05/2023
14391	2023	All other reportable incidents	47265	Dual-chamber implantable pacemaker, rate-responsive	CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A0908	Incorrect, Inadequate or Imprecise Result or Readings		
							F26	No Health Consequences or Impact	A0908	Incorrect, Inadequate or Imprecise Result or Readings		
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A0908	Incorrect, Inadequate or Imprecise Result or Readings		
							F26	No Health Consequences or Impact	A0908	Incorrect, Inadequate or Imprecise Result or Readings		
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A0908	Incorrect, Inadequate or Imprecise Result or Readings		
							F26	No Health Consequences or Impact	A0908	Incorrect, Inadequate or Imprecise Result or Readings		
14392	2023	All other reportable incidents	47265	Dual-chamber implantable pacemaker, rate-responsive	CT141	Electrodes/Leads and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A090801	Signal Artifact/Noise	11/07/2023	11/07/2023
							F1908	Prolonged Surgery	A090801	Signal Artifact/Noise	11/07/2023	11/07/2023
					CT145	Cardiovascular devices	E2403	No Clinical Signs, Symptoms or Conditions	A090801	Signal Artifact/Noise	11/07/2023	11/07/2023
							F1908	Prolonged Surgery	A090801	Signal Artifact/Noise	11/07/2023	11/07/2023
					CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A090801	Signal Artifact/Noise	11/07/2023	11/07/2023
							F1908	Prolonged Surgery	A090801	Signal Artifact/Noise	11/07/2023	11/07/2023
					CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A090801	Signal Artifact/Noise	11/07/2023	11/07/2023
							F1908	Prolonged Surgery	A090801	Signal Artifact/Noise	11/07/2023	11/07/2023
					CT233	Surgical	E2403	No Clinical Signs, Symptoms or Conditions	A090801	Signal Artifact/Noise	11/07/2023	11/07/2023
							F1908	Prolonged Surgery	A090801	Signal Artifact/Noise	11/07/2023	11/07/2023
					CT2406	Implantable	E2403	No Clinical Signs, Symptoms or Conditions	A090801	Signal Artifact/Noise	11/07/2023	11/07/2023
							F1908	Prolonged Surgery	A090801	Signal Artifact/Noise	11/07/2023	11/07/2023
					CT2407	Active implantable	E2403	No Clinical Signs, Symptoms or Conditions	A090801	Signal Artifact/Noise	11/07/2023	11/07/2023
							F1908	Prolonged Surgery	A090801	Signal Artifact/Noise	11/07/2023	11/07/2023
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A090801	Signal Artifact/Noise	11/07/2023	11/07/2023
							F1908	Prolonged Surgery	A090801	Signal Artifact/Noise	11/07/2023	11/07/2023
					CT2727	Active implant devices/ancillaries	E2403	No Clinical Signs, Symptoms or Conditions	A090801	Signal Artifact/Noise	11/07/2023	11/07/2023
							F1908	Prolonged Surgery	A090801	Signal Artifact/Noise	11/07/2023	11/07/2023
					CT2750	Implantable cardiac electrical stimulators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A090801	Signal Artifact/Noise	11/07/2023	11/07/2023
							F1908	Prolonged Surgery	A090801	Signal Artifact/Noise	11/07/2023	11/07/2023
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A090801	Signal Artifact/Noise	11/07/2023	11/07/2023
							F1908	Prolonged Surgery	A090801	Signal Artifact/Noise	11/07/2023	11/07/2023
					CT2985	Active	E2403	No Clinical Signs, Symptoms or Conditions	A090801	Signal Artifact/Noise	11/07/2023	11/07/2023
							F1908	Prolonged Surgery	A090801	Signal Artifact/Noise	11/07/2023	11/07/2023
					CT3001	Implantable metal	E2403	No Clinical Signs, Symptoms or Conditions	A090801	Signal Artifact/Noise	11/07/2023	11/07/2023
							F1908	Prolonged Surgery	A090801	Signal Artifact/Noise	11/07/2023	11/07/2023
					CT3002	Implantable synthetic polymer	E2403	No Clinical Signs, Symptoms or Conditions	A090801	Signal Artifact/Noise	11/07/2023	11/07/2023
							F1908	Prolonged Surgery	A090801	Signal Artifact/Noise	11/07/2023	11/07/2023
					CT3006	Implantable inorganic material	E2403	No Clinical Signs, Symptoms or Conditions	A090801	Signal Artifact/Noise	11/07/2023	11/07/2023
							F1908	Prolonged Surgery	A090801	Signal Artifact/Noise	11/07/2023	11/07/2023
					CT314	Electrically powered	E2403	No Clinical Signs, Symptoms or Conditions	A090801	Signal Artifact/Noise	11/07/2023	11/07/2023
							F1908	Prolonged Surgery	A090801	Signal Artifact/Noise	11/07/2023	11/07/2023
					CT321	Long-term surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A090801	Signal Artifact/Noise	11/07/2023	11/07/2023
							F1908	Prolonged Surgery	A090801	Signal Artifact/Noise	11/07/2023	11/07/2023
					CT449	Generators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A090801	Signal Artifact/Noise	11/07/2023	11/07/2023
							F1908	Prolonged Surgery	A090801	Signal Artifact/Noise	11/07/2023	11/07/2023
					CT981	Single use	E2403	No Clinical Signs, Symptoms or Conditions	A090801	Signal Artifact/Noise	11/07/2023	11/07/2023
							F1908	Prolonged Surgery	A090801	Signal Artifact/Noise	11/07/2023	11/07/2023
					CT983	Surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A090801	Signal Artifact/Noise	11/07/2023	11/07/2023
							F1908	Prolonged Surgery	A090801	Signal Artifact/Noise	11/07/2023	11/07/2023
14393	2023		47265	Dual-chamber implantable pacemaker, rate-responsive	CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	31/10/2023	31/10/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	31/10/2023	31/10/2023
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	31/10/2023	31/10/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	31/10/2023	31/10/2023
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	31/10/2023	31/10/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	31/10/2023	31/10/2023
14394	2023	All other reportable incidents	17882	Physiologic-monitoring defibrillation system	CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070908	Failure to Sense		
									A070912	Incorrect Interpretation of Signal		
							F26	No Health Consequences or Impact	A070908	Failure to Sense		
									A070912	Incorrect Interpretation of Signal		
14395	2023		47265	Dual-chamber implantable pacemaker, rate-responsive	CT141	Electrodes/Leads and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	05/09/2023	05/09/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	05/09/2023	05/09/2023
					CT145	Cardiovascular devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	05/09/2023	05/09/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	05/09/2023	05/09/2023
					CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	05/09/2023	05/09/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	05/09/2023	05/09/2023
					CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	05/09/2023	05/09/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	05/09/2023	05/09/2023
					CT233	Surgical	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	05/09/2023	05/09/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	05/09/2023	05/09/2023
					CT2406	Implantable	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	05/09/2023	05/09/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	05/09/2023	05/09/2023
					CT2407	Active implantable	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	05/09/2023	05/09/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	05/09/2023	05/09/2023
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	05/09/2023	05/09/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	05/09/2023	05/09/2023
					CT2727	Active implant devices/ancillaries	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	05/09/2023	05/09/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	05/09/2023	05/09/2023
					CT2750	Implantable cardiac electrical stimulators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	05/09/2023	05/09/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	05/09/2023	05/09/2023
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	05/09/2023	05/09/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	05/09/2023	05/09/2023
					CT2985	Active	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	05/09/2023	05/09/2023

							F1901	Additional Surgery	A24	No Apparent Device Problem	05/09/2023	05/09/2023
					CT3001	Implantable metal	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	05/09/2023	05/09/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	05/09/2023	05/09/2023
					CT3002	Implantable synthetic polymer	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	05/09/2023	05/09/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	05/09/2023	05/09/2023
					CT3006	Implantable inorganic material	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	05/09/2023	05/09/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	05/09/2023	05/09/2023
					CT314	Electrically powered	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	05/09/2023	05/09/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	05/09/2023	05/09/2023
					CT321	Long-term surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	05/09/2023	05/09/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	05/09/2023	05/09/2023
					CT449	Generators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	05/09/2023	05/09/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	05/09/2023	05/09/2023
					CT981	Single use	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	05/09/2023	05/09/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	05/09/2023	05/09/2023
					CT983	Surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	05/09/2023	05/09/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	05/09/2023	05/09/2023
14396	2023	All other reportable incidents	17882	Physiologic-monitoring defibrillation system	CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A07	Electrical /Electronic Property Problem		
							F27	Problem Identified During Non-Clinical Procedure	A07	Electrical /Electronic Property Problem		
14397	2023	All other reportable incidents	47267	Single-chamber implantable pacemaker, rate-responsive	CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
14398	2023	All other reportable incidents	35852	Single-chamber implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	08/12/2023	08/12/2023
									A0908	Incorrect, Inadequate or Imprecise Result or Readings	08/12/2023	08/12/2023
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	08/12/2023	08/12/2023
									A0908	Incorrect, Inadequate or Imprecise Result or Readings	08/12/2023	08/12/2023
							F1903	Device Explantation	A070504	Premature Discharge of Battery	08/12/2023	08/12/2023
									A0908	Incorrect, Inadequate or Imprecise Result or Readings	08/12/2023	08/12/2023
14399	2023	All other reportable incidents	35223	Endocardial/interventricular septal pacing lead	CT1684	Cardiac pacemakers and associated devices	E210401	Shock from Patient Lead(s)	A070101	Failure to Capture		
									A070910	Undersensing		
							F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A070101	Failure to Capture		
									A070910	Undersensing		
							F1203	Life Threatening Illness or Injury	A070101	Failure to Capture		
									A070910	Undersensing		
							F23	Unexpected Medical Intervention	A070101	Failure to Capture		
									A070910	Undersensing		
							F2303	Medication Required	A070101	Failure to Capture		
									A070910	Undersensing		
14400	2023	All other reportable incidents	17882	Physiologic-monitoring defibrillation system	CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A07	Electrical /Electronic Property Problem		
							F27	Problem Identified During Non-Clinical Procedure	A07	Electrical /Electronic Property Problem		
14401	2023	All other reportable incidents	37265	Dual-chamber implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A160602	Premature End-of-Life Indicator	07/11/2023	07/11/2023
							F1903	Device Explantation	A160602	Premature End-of-Life Indicator	07/11/2023	07/11/2023
14402	2023	All other reportable incidents	35852	Single-chamber implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	01/11/2023	01/11/2023
									A160601	Premature Elective Replacement Indicator	01/11/2023	01/11/2023
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	01/11/2023	01/11/2023
									A160601	Premature Elective Replacement Indicator	01/11/2023	01/11/2023
							F1903	Device Explantation	A070504	Premature Discharge of Battery	01/11/2023	01/11/2023
									A160601	Premature Elective Replacement Indicator	01/11/2023	01/11/2023
14403	2023		47265	Dual-chamber implantable pacemaker, rate-responsive	CT141	Electrodes/Leads and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	17/10/2023	17/10/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	17/10/2023	17/10/2023
					CT145	Cardiovascular devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	17/10/2023	17/10/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	17/10/2023	17/10/2023
					CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	17/10/2023	17/10/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	17/10/2023	17/10/2023
					CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	17/10/2023	17/10/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	17/10/2023	17/10/2023
					CT233	Surgical	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	17/10/2023	17/10/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	17/10/2023	17/10/2023
					CT2406	Implantable	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	17/10/2023	17/10/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	17/10/2023	17/10/2023
					CT2407	Active implantable	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	17/10/2023	17/10/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	17/10/2023	17/10/2023
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	17/10/2023	17/10/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	17/10/2023	17/10/2023
					CT2727	Active implant devices/ancillaries	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	17/10/2023	17/10/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	17/10/2023	17/10/2023
					CT2750	Implantable cardiac electrical stimulators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	17/10/2023	17/10/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	17/10/2023	17/10/2023
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	17/10/2023	17/10/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	17/10/2023	17/10/2023
					CT2985	Active	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	17/10/2023	17/10/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	17/10/2023	17/10/2023
					CT3001	Implantable metal	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	17/10/2023	17/10/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	17/10/2023	17/10/2023
					CT3002	Implantable synthetic polymer	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	17/10/2023	17/10/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	17/10/2023	17/10/2023
					CT3006	Implantable inorganic material	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	17/10/2023	17/10/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	17/10/2023	17/10/2023
					CT314	Electrically powered	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	17/10/2023	17/10/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	17/10/2023	17/10/2023
					CT321	Long-term surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	17/10/2023	17/10/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	17/10/2023	17/10/2023

					CT449	Generators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	17/10/2023	17/10/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	17/10/2023	17/10/2023
					CT981	Single use	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	17/10/2023	17/10/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	17/10/2023	17/10/2023
					CT983	Surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	17/10/2023	17/10/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	17/10/2023	17/10/2023
14404	2023	All other reportable incidents	47267	Single-chamber implantable pacemaker, rate-responsive	CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
14405	2023	All other reportable incidents	37265	Dual-chamber implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A072201	High Impedance		
							F26	No Health Consequences or Impact	A072201	High Impedance		
14406	2023		47265	Dual-chamber implantable pacemaker, rate-responsive	CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	31/10/2023	31/10/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	31/10/2023	31/10/2023
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	31/10/2023	31/10/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	31/10/2023	31/10/2023
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	31/10/2023	31/10/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	31/10/2023	31/10/2023
14407	2023	All other reportable incidents	17882	Physiologic-monitoring defibrillation system	CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070803	Failure to Power Up		
							F27	Problem Identified During Non-Clinical Procedure	A070803	Failure to Power Up		
14408	2023		47265	Dual-chamber implantable pacemaker, rate-responsive	CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	10/11/2023	10/11/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	10/11/2023	10/11/2023
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	10/11/2023	10/11/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	10/11/2023	10/11/2023
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	10/11/2023	10/11/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	10/11/2023	10/11/2023
14409	2023		47265	Dual-chamber implantable pacemaker, rate-responsive	CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	10/11/2023	10/11/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	10/11/2023	10/11/2023
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	10/11/2023	10/11/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	10/11/2023	10/11/2023
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	10/11/2023	10/11/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	10/11/2023	10/11/2023
14410	2023		47265	Dual-chamber implantable pacemaker, rate-responsive	CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	10/11/2023	10/11/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	10/11/2023	10/11/2023
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	10/11/2023	10/11/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	10/11/2023	10/11/2023
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	10/11/2023	10/11/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	10/11/2023	10/11/2023
14411	2023	All other reportable incidents	47205	Cardiac pulse generator programmer	CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A150105	Key or Button Unresponsive/not Working		
							F26	No Health Consequences or Impact	A150105	Key or Button Unresponsive/not Working		
14412	2023	All other reportable incidents	47265	Dual-chamber implantable pacemaker, rate-responsive	CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	02/04/2024	02/04/2024
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	02/04/2024	02/04/2024
							F1903	Device Explantation	A070504	Premature Discharge of Battery	02/04/2024	02/04/2024
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	02/04/2024	02/04/2024
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	02/04/2024	02/04/2024
							F1903	Device Explantation	A070504	Premature Discharge of Battery	02/04/2024	02/04/2024
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	02/04/2024	02/04/2024
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	02/04/2024	02/04/2024
							F1903	Device Explantation	A070504	Premature Discharge of Battery	02/04/2024	02/04/2024
14413	2023	All other reportable incidents	17882	Physiologic-monitoring defibrillation system	CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A07	Electrical /Electronic Property Problem		
							F27	Problem Identified During Non-Clinical Procedure	A07	Electrical /Electronic Property Problem		
14414	2023	All other reportable incidents	35223	Endocardial/interventricular septal pacing lead	CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A0401	Break	04/09/2023	04/09/2023
									A070909	Oversensing	04/09/2023	04/09/2023
							F1901	Additional Surgery	A0401	Break	04/09/2023	04/09/2023
									A070909	Oversensing	04/09/2023	04/09/2023
14415	2023	All other reportable incidents	47265	Dual-chamber implantable pacemaker, rate-responsive	CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem		
							F1901	Additional Surgery	A24	No Apparent Device Problem		
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem		
							F1901	Additional Surgery	A24	No Apparent Device Problem		
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem		
							F1901	Additional Surgery	A24	No Apparent Device Problem		
14416	2023	All other reportable incidents	47265	Dual-chamber implantable pacemaker, rate-responsive	CT141	Electrodes/Leads and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070909	Oversensing		
									A070910	Undersensing		
									A090801	Signal Artifact/Noise		
									A1301	Failure to Read Input Signal		
							F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A070909	Oversensing		
									A070910	Undersensing		
									A090801	Signal Artifact/Noise		
									A1301	Failure to Read Input Signal		
					CT145	Cardiovascular devices	E2403	No Clinical Signs, Symptoms or Conditions	A070909	Oversensing		
									A070910	Undersensing		
									A090801	Signal Artifact/Noise		
									A1301	Failure to Read Input Signal		
							F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A070909	Oversensing		
									A070910	Undersensing		
									A090801	Signal Artifact/Noise		
									A1301	Failure to Read Input Signal		
					CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070909	Oversensing		
									A070910	Undersensing		
									A090801	Signal Artifact/Noise		

									A1301	Failure to Read Input Signal		
						F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A070909	Oversensing			
								A070910	Undersensing			
								A090801	Signal Artifact/Noise			
								A1301	Failure to Read Input Signal			
					CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A070909	Oversensing		
									A070910	Undersensing		
									A090801	Signal Artifact/Noise		
									A1301	Failure to Read Input Signal		
							F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A070909	Oversensing		
									A070910	Undersensing		
									A090801	Signal Artifact/Noise		
									A1301	Failure to Read Input Signal		
					CT233	Surgical	E2403	No Clinical Signs, Symptoms or Conditions	A070909	Oversensing		
									A070910	Undersensing		
									A090801	Signal Artifact/Noise		
									A1301	Failure to Read Input Signal		
							F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A070909	Oversensing		
									A070910	Undersensing		
									A090801	Signal Artifact/Noise		
									A1301	Failure to Read Input Signal		
					CT2406	Implantable	E2403	No Clinical Signs, Symptoms or Conditions	A070909	Oversensing		
									A070910	Undersensing		
									A090801	Signal Artifact/Noise		
									A1301	Failure to Read Input Signal		
							F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A070909	Oversensing		
									A070910	Undersensing		
									A090801	Signal Artifact/Noise		
									A1301	Failure to Read Input Signal		
					CT2407	Active implantable	E2403	No Clinical Signs, Symptoms or Conditions	A070909	Oversensing		
									A070910	Undersensing		
									A090801	Signal Artifact/Noise		
									A1301	Failure to Read Input Signal		
							F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A070909	Oversensing		
									A070910	Undersensing		
									A090801	Signal Artifact/Noise		
									A1301	Failure to Read Input Signal		
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A070909	Oversensing		
									A070910	Undersensing		
									A090801	Signal Artifact/Noise		
									A1301	Failure to Read Input Signal		
							F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A070909	Oversensing		
									A070910	Undersensing		
									A090801	Signal Artifact/Noise		
									A1301	Failure to Read Input Signal		
					CT2727	Active implant devices/ancillaries	E2403	No Clinical Signs, Symptoms or Conditions	A070909	Oversensing		
									A070910	Undersensing		
									A090801	Signal Artifact/Noise		
									A1301	Failure to Read Input Signal		
							F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A070909	Oversensing		
									A070910	Undersensing		
									A090801	Signal Artifact/Noise		
									A1301	Failure to Read Input Signal		
					CT2750	Implantable cardiac electrical stimulators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070909	Oversensing		
									A070910	Undersensing		
									A090801	Signal Artifact/Noise		
									A1301	Failure to Read Input Signal		
							F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A070909	Oversensing		
									A070910	Undersensing		
									A090801	Signal Artifact/Noise		
									A1301	Failure to Read Input Signal		
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A070909	Oversensing		
									A070910	Undersensing		
									A090801	Signal Artifact/Noise		
									A1301	Failure to Read Input Signal		
							F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A070909	Oversensing		
									A070910	Undersensing		
									A090801	Signal Artifact/Noise		
									A1301	Failure to Read Input Signal		
					CT2985	Active	E2403	No Clinical Signs, Symptoms or Conditions	A070909	Oversensing		
									A070910	Undersensing		
									A090801	Signal Artifact/Noise		
									A1301	Failure to Read Input Signal		
							F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A070909	Oversensing		
									A070910	Undersensing		
									A090801	Signal Artifact/Noise		
									A1301	Failure to Read Input Signal		
					CT3001	Implantable metal	E2403	No Clinical Signs, Symptoms or Conditions	A070909	Oversensing		
									A070910	Undersensing		
									A090801	Signal Artifact/Noise		
									A1301	Failure to Read Input Signal		
							F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A070909	Oversensing		
									A070910	Undersensing		
									A090801	Signal Artifact/Noise		
									A1301	Failure to Read Input Signal		
					CT3002	Implantable synthetic polymer	E2403	No Clinical Signs, Symptoms or Conditions	A070909	Oversensing		
									A070910	Undersensing		
									A090801	Signal Artifact/Noise		

									A1301	Failure to Read Input Signal		
							F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A070909	Oversensing		
									A070910	Undersensing		
									A090801	Signal Artifact/Noise		
									A1301	Failure to Read Input Signal		
					CT3006	Implantable inorganic material	E2403	No Clinical Signs, Symptoms or Conditions	A070909	Oversensing		
									A070910	Undersensing		
									A090801	Signal Artifact/Noise		
									A1301	Failure to Read Input Signal		
							F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A070909	Oversensing		
									A070910	Undersensing		
									A090801	Signal Artifact/Noise		
									A1301	Failure to Read Input Signal		
					CT314	Electrically powered	E2403	No Clinical Signs, Symptoms or Conditions	A070909	Oversensing		
									A070910	Undersensing		
									A090801	Signal Artifact/Noise		
									A1301	Failure to Read Input Signal		
							F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A070909	Oversensing		
									A070910	Undersensing		
									A090801	Signal Artifact/Noise		
									A1301	Failure to Read Input Signal		
					CT321	Long-term surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A070909	Oversensing		
									A070910	Undersensing		
									A090801	Signal Artifact/Noise		
									A1301	Failure to Read Input Signal		
							F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A070909	Oversensing		
									A070910	Undersensing		
									A090801	Signal Artifact/Noise		
									A1301	Failure to Read Input Signal		
					CT449	Generators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070909	Oversensing		
									A070910	Undersensing		
									A090801	Signal Artifact/Noise		
									A1301	Failure to Read Input Signal		
							F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A070909	Oversensing		
									A070910	Undersensing		
									A090801	Signal Artifact/Noise		
									A1301	Failure to Read Input Signal		
					CT981	Single use	E2403	No Clinical Signs, Symptoms or Conditions	A070909	Oversensing		
									A070910	Undersensing		
									A090801	Signal Artifact/Noise		
									A1301	Failure to Read Input Signal		
							F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A070909	Oversensing		
									A070910	Undersensing		
									A090801	Signal Artifact/Noise		
									A1301	Failure to Read Input Signal		
					CT983	Surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A070909	Oversensing		
									A070910	Undersensing		
									A090801	Signal Artifact/Noise		
									A1301	Failure to Read Input Signal		
							F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A070909	Oversensing		
									A070910	Undersensing		
									A090801	Signal Artifact/Noise		
									A1301	Failure to Read Input Signal		
14417	2023	All other reportable incidents	37265	Dual-chamber implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E2104	Electric Shock	A071302	Inappropriate/Inadequate Shock/Stimulation	23/10/2023	23/10/2023
									A090801	Signal Artifact/Noise	23/10/2023	23/10/2023
									A1301	Failure to Read Input Signal	23/10/2023	23/10/2023
							F08	Hospitalization or Prolonged Hospitalization	A071302	Inappropriate/Inadequate Shock/Stimulation	23/10/2023	23/10/2023
									A090801	Signal Artifact/Noise	23/10/2023	23/10/2023
									A1301	Failure to Read Input Signal	23/10/2023	23/10/2023
							F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A071302	Inappropriate/Inadequate Shock/Stimulation	23/10/2023	23/10/2023
									A090801	Signal Artifact/Noise	23/10/2023	23/10/2023
									A1301	Failure to Read Input Signal	23/10/2023	23/10/2023
							F1905	Device Revision or Replacement	A071302	Inappropriate/Inadequate Shock/Stimulation	23/10/2023	23/10/2023
									A090801	Signal Artifact/Noise	23/10/2023	23/10/2023
									A1301	Failure to Read Input Signal	23/10/2023	23/10/2023
14418	2023	All other reportable incidents			CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A05	Mechanical Problem	07/12/2023	07/12/2023
									A0705	Battery Problem	07/12/2023	07/12/2023
									A070504	Premature Discharge of Battery	07/12/2023	07/12/2023
							F08	Hospitalization or Prolonged Hospitalization	A05	Mechanical Problem	07/12/2023	07/12/2023
									A0705	Battery Problem	07/12/2023	07/12/2023
									A070504	Premature Discharge of Battery	07/12/2023	07/12/2023
							F1905	Device Revision or Replacement	A05	Mechanical Problem	07/12/2023	07/12/2023
									A0705	Battery Problem	07/12/2023	07/12/2023
									A070504	Premature Discharge of Battery	07/12/2023	07/12/2023
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A05	Mechanical Problem	07/12/2023	07/12/2023
									A0705	Battery Problem	07/12/2023	07/12/2023
									A070504	Premature Discharge of Battery	07/12/2023	07/12/2023
							F08	Hospitalization or Prolonged Hospitalization	A05	Mechanical Problem	07/12/2023	07/12/2023
									A0705	Battery Problem	07/12/2023	07/12/2023
									A070504	Premature Discharge of Battery	07/12/2023	07/12/2023
							F1905	Device Revision or Replacement	A05	Mechanical Problem	07/12/2023	07/12/2023
									A0705	Battery Problem	07/12/2023	07/12/2023
									A070504	Premature Discharge of Battery	07/12/2023	07/12/2023
14419	2023		17882	Physiologic-monitoring defibrillation system	CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A090810	Low Test Results		
							F27	Problem Identified During Non-Clinical Procedure	A090810	Low Test Results		
14420	2023	All other reportable incidents	47265	Dual-chamber implantable pacemaker, rate-responsive	CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A05	Mechanical Problem	07/11/2023	07/11/2023
							F08	Hospitalization or Prolonged Hospitalization	A05	Mechanical Problem	07/11/2023	07/11/2023
							F12	Serious Injury/ Illness/ Impairment	A05	Mechanical Problem	07/11/2023	07/11/2023

							F1905	Device Revision or Replacement	A05	Mechanical Problem	07/11/2023	07/11/2023
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A05	Mechanical Problem	07/11/2023	07/11/2023
							F08	Hospitalization or Prolonged Hospitalization	A05	Mechanical Problem	07/11/2023	07/11/2023
							F12	Serious Injury/ Illness/ Impairment	A05	Mechanical Problem	07/11/2023	07/11/2023
							F1905	Device Revision or Replacement	A05	Mechanical Problem	07/11/2023	07/11/2023
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A05	Mechanical Problem	07/11/2023	07/11/2023
							F08	Hospitalization or Prolonged Hospitalization	A05	Mechanical Problem	07/11/2023	07/11/2023
							F12	Serious Injury/ Illness/ Impairment	A05	Mechanical Problem	07/11/2023	07/11/2023
							F1905	Device Revision or Replacement	A05	Mechanical Problem	07/11/2023	07/11/2023
14421	2023		47267	Single-chamber implantable pacemaker, rate-responsive	CT141	Electrodes/Leads and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A0203	Defective Device		
									A071102	Failure to Interrogate		
									A071204	No Pacing		
							F19	Surgical Intervention	A0203	Defective Device		
									A071102	Failure to Interrogate		
									A071204	No Pacing		
							F1905	Device Revision or Replacement	A0203	Defective Device		
									A071102	Failure to Interrogate		
									A071204	No Pacing		
					CT145	Cardiovascular devices	E2403	No Clinical Signs, Symptoms or Conditions	A0203	Defective Device		
									A071102	Failure to Interrogate		
									A071204	No Pacing		
							F19	Surgical Intervention	A0203	Defective Device		
									A071102	Failure to Interrogate		
									A071204	No Pacing		
							F1905	Device Revision or Replacement	A0203	Defective Device		
									A071102	Failure to Interrogate		
									A071204	No Pacing		
					CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A0203	Defective Device		
									A071102	Failure to Interrogate		
									A071204	No Pacing		
							F19	Surgical Intervention	A0203	Defective Device		
									A071102	Failure to Interrogate		
									A071204	No Pacing		
							F1905	Device Revision or Replacement	A0203	Defective Device		
									A071102	Failure to Interrogate		
									A071204	No Pacing		
					CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A0203	Defective Device		
									A071102	Failure to Interrogate		
									A071204	No Pacing		
							F19	Surgical Intervention	A0203	Defective Device		
									A071102	Failure to Interrogate		
									A071204	No Pacing		
							F1905	Device Revision or Replacement	A0203	Defective Device		
									A071102	Failure to Interrogate		
									A071204	No Pacing		
					CT233	Surgical	E2403	No Clinical Signs, Symptoms or Conditions	A0203	Defective Device		
									A071102	Failure to Interrogate		
									A071204	No Pacing		
							F19	Surgical Intervention	A0203	Defective Device		
									A071102	Failure to Interrogate		
									A071204	No Pacing		
							F1905	Device Revision or Replacement	A0203	Defective Device		
									A071102	Failure to Interrogate		
									A071204	No Pacing		
					CT2406	Implantable	E2403	No Clinical Signs, Symptoms or Conditions	A0203	Defective Device		
									A071102	Failure to Interrogate		
									A071204	No Pacing		
							F19	Surgical Intervention	A0203	Defective Device		
									A071102	Failure to Interrogate		
									A071204	No Pacing		
							F1905	Device Revision or Replacement	A0203	Defective Device		
									A071102	Failure to Interrogate		
									A071204	No Pacing		
					CT2407	Active implantable	E2403	No Clinical Signs, Symptoms or Conditions	A0203	Defective Device		
									A071102	Failure to Interrogate		
									A071204	No Pacing		
							F19	Surgical Intervention	A0203	Defective Device		
									A071102	Failure to Interrogate		
									A071204	No Pacing		
							F1905	Device Revision or Replacement	A0203	Defective Device		
									A071102	Failure to Interrogate		
									A071204	No Pacing		
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A0203	Defective Device		
									A071102	Failure to Interrogate		
									A071204	No Pacing		
							F19	Surgical Intervention	A0203	Defective Device		
									A071102	Failure to Interrogate		
									A071204	No Pacing		
							F1905	Device Revision or Replacement	A0203	Defective Device		
									A071102	Failure to Interrogate		
									A071204	No Pacing		
					CT2727	Active implant devices/ancillaries	E2403	No Clinical Signs, Symptoms or Conditions	A0203	Defective Device		
									A071102	Failure to Interrogate		
									A071204	No Pacing		
							F19	Surgical Intervention	A0203	Defective Device		
									A071102	Failure to Interrogate		
									A071204	No Pacing		
							F1905	Device Revision or Replacement	A0203	Defective Device		

								A071102	Failure to Interrogate		
								A071204	No Pacing		
					CT2750	Implantable cardiac electrical stimulators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A0203	Defective Device		
									A071102	Failure to Interrogate		
									A071204	No Pacing		
							F19	Surgical Intervention	A0203	Defective Device		
									A071102	Failure to Interrogate		
									A071204	No Pacing		
							F1905	Device Revision or Replacement	A0203	Defective Device		
									A071102	Failure to Interrogate		
									A071204	No Pacing		
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A0203	Defective Device		
									A071102	Failure to Interrogate		
									A071204	No Pacing		
							F19	Surgical Intervention	A0203	Defective Device		
									A071102	Failure to Interrogate		
									A071204	No Pacing		
							F1905	Device Revision or Replacement	A0203	Defective Device		
									A071102	Failure to Interrogate		
									A071204	No Pacing		
					CT2985	Active	E2403	No Clinical Signs, Symptoms or Conditions	A0203	Defective Device		
									A071102	Failure to Interrogate		
									A071204	No Pacing		
							F19	Surgical Intervention	A0203	Defective Device		
									A071102	Failure to Interrogate		
									A071204	No Pacing		
							F1905	Device Revision or Replacement	A0203	Defective Device		
									A071102	Failure to Interrogate		
									A071204	No Pacing		
					CT3001	Implantable metal	E2403	No Clinical Signs, Symptoms or Conditions	A0203	Defective Device		
									A071102	Failure to Interrogate		
									A071204	No Pacing		
							F19	Surgical Intervention	A0203	Defective Device		
									A071102	Failure to Interrogate		
									A071204	No Pacing		
							F1905	Device Revision or Replacement	A0203	Defective Device		
									A071102	Failure to Interrogate		
									A071204	No Pacing		
					CT3002	Implantable synthetic polymer	E2403	No Clinical Signs, Symptoms or Conditions	A0203	Defective Device		
									A071102	Failure to Interrogate		
									A071204	No Pacing		
							F19	Surgical Intervention	A0203	Defective Device		
									A071102	Failure to Interrogate		
									A071204	No Pacing		
							F1905	Device Revision or Replacement	A0203	Defective Device		
									A071102	Failure to Interrogate		
									A071204	No Pacing		
					CT3006	Implantable inorganic material	E2403	No Clinical Signs, Symptoms or Conditions	A0203	Defective Device		
									A071102	Failure to Interrogate		
									A071204	No Pacing		
							F19	Surgical Intervention	A0203	Defective Device		
									A071102	Failure to Interrogate		
									A071204	No Pacing		
							F1905	Device Revision or Replacement	A0203	Defective Device		
									A071102	Failure to Interrogate		
									A071204	No Pacing		
					CT314	Electrically powered	E2403	No Clinical Signs, Symptoms or Conditions	A0203	Defective Device		
									A071102	Failure to Interrogate		
									A071204	No Pacing		
							F19	Surgical Intervention	A0203	Defective Device		
									A071102	Failure to Interrogate		
									A071204	No Pacing		
							F1905	Device Revision or Replacement	A0203	Defective Device		
									A071102	Failure to Interrogate		
									A071204	No Pacing		
					CT321	Long-term surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A0203	Defective Device		
									A071102	Failure to Interrogate		
									A071204	No Pacing		
							F19	Surgical Intervention	A0203	Defective Device		
									A071102	Failure to Interrogate		
									A071204	No Pacing		
							F1905	Device Revision or Replacement	A0203	Defective Device		
									A071102	Failure to Interrogate		
									A071204	No Pacing		
					CT449	Generators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A0203	Defective Device		
									A071102	Failure to Interrogate		
									A071204	No Pacing		
							F19	Surgical Intervention	A0203	Defective Device		
									A071102	Failure to Interrogate		
									A071204	No Pacing		
							F1905	Device Revision or Replacement	A0203	Defective Device		
									A071102	Failure to Interrogate		
									A071204	No Pacing		
					CT981	Single use	E2403	No Clinical Signs, Symptoms or Conditions	A0203	Defective Device		
									A071102	Failure to Interrogate		
									A071204	No Pacing		
							F19	Surgical Intervention	A0203	Defective Device		
									A071102	Failure to Interrogate		

								A071204	No Pacing			
						F1905	Device Revision or Replacement	A0203	Defective Device			
								A071102	Failure to Interrogate			
								A071204	No Pacing			
					CT983	Surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A0203	Defective Device		
									A071102	Failure to Interrogate		
									A071204	No Pacing		
							F19	Surgical Intervention	A0203	Defective Device		
									A071102	Failure to Interrogate		
									A071204	No Pacing		
							F1905	Device Revision or Replacement	A0203	Defective Device		
									A071102	Failure to Interrogate		
									A071204	No Pacing		
14422	2023	All other reportable incidents	17882	Physiologic-monitoring defibrillation system	CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A0719	Unexpected Shutdown		
							F26	No Health Consequences or Impact	A0719	Unexpected Shutdown		
14423	2023		47265	Dual-chamber implantable pacemaker, rate-responsive	CT141	Electrodes/Leads and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	27/10/2023	27/10/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	27/10/2023	27/10/2023
					CT145	Cardiovascular devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	27/10/2023	27/10/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	27/10/2023	27/10/2023
					CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	27/10/2023	27/10/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	27/10/2023	27/10/2023
					CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	27/10/2023	27/10/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	27/10/2023	27/10/2023
					CT233	Surgical	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	27/10/2023	27/10/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	27/10/2023	27/10/2023
					CT2406	Implantable	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	27/10/2023	27/10/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	27/10/2023	27/10/2023
					CT2407	Active implantable	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	27/10/2023	27/10/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	27/10/2023	27/10/2023
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	27/10/2023	27/10/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	27/10/2023	27/10/2023
					CT2727	Active implant devices/ancillaries	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	27/10/2023	27/10/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	27/10/2023	27/10/2023
					CT2750	Implantable cardiac electrical stimulators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	27/10/2023	27/10/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	27/10/2023	27/10/2023
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	27/10/2023	27/10/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	27/10/2023	27/10/2023
					CT2985	Active	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	27/10/2023	27/10/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	27/10/2023	27/10/2023
					CT3001	Implantable metal	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	27/10/2023	27/10/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	27/10/2023	27/10/2023
					CT3002	Implantable synthetic polymer	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	27/10/2023	27/10/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	27/10/2023	27/10/2023
					CT3006	Implantable inorganic material	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	27/10/2023	27/10/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	27/10/2023	27/10/2023
					CT314	Electrically powered	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	27/10/2023	27/10/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	27/10/2023	27/10/2023
					CT321	Long-term surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	27/10/2023	27/10/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	27/10/2023	27/10/2023
					CT449	Generators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	27/10/2023	27/10/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	27/10/2023	27/10/2023
					CT981	Single use	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	27/10/2023	27/10/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	27/10/2023	27/10/2023
					CT983	Surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	27/10/2023	27/10/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	27/10/2023	27/10/2023
14424	2023		47265	Dual-chamber implantable pacemaker, rate-responsive	CT141	Electrodes/Leads and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	07/11/2023	07/11/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	07/11/2023	07/11/2023
					CT145	Cardiovascular devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	07/11/2023	07/11/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	07/11/2023	07/11/2023
					CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	07/11/2023	07/11/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	07/11/2023	07/11/2023
					CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	07/11/2023	07/11/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	07/11/2023	07/11/2023
					CT233	Surgical	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	07/11/2023	07/11/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	07/11/2023	07/11/2023
					CT2406	Implantable	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	07/11/2023	07/11/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	07/11/2023	07/11/2023
					CT2407	Active implantable	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	07/11/2023	07/11/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	07/11/2023	07/11/2023
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	07/11/2023	07/11/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	07/11/2023	07/11/2023
					CT2727	Active implant devices/ancillaries	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	07/11/2023	07/11/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	07/11/2023	07/11/2023
					CT2750	Implantable cardiac electrical stimulators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	07/11/2023	07/11/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	07/11/2023	07/11/2023
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	07/11/2023	07/11/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	07/11/2023	07/11/2023
					CT2985	Active	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	07/11/2023	07/11/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	07/11/2023	07/11/2023
					CT3001	Implantable metal	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	07/11/2023	07/11/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	07/11/2023	07/11/2023
					CT3002	Implantable synthetic polymer	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	07/11/2023	07/11/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	07/11/2023	07/11/2023
					CT3006	Implantable inorganic material	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	07/11/2023	07/11/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	07/11/2023	07/11/2023
					CT314	Electrically powered	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	07/11/2023	07/11/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	07/11/2023	07/11/2023
					CT321	Long-term surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	07/11/2023	07/11/2023

							F1901	Additional Surgery	A24	No Apparent Device Problem	07/11/2023	07/11/2023
					CT449	Generators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	07/11/2023	07/11/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	07/11/2023	07/11/2023
					CT981	Single use	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	07/11/2023	07/11/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	07/11/2023	07/11/2023
					CT983	Surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	07/11/2023	07/11/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	07/11/2023	07/11/2023
14425	2023	All other reportable incidents	37265	Dual-chamber implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A1304	Telemetry Discrepancy	19/10/2023	19/10/2023
							F08	Hospitalization or Prolonged Hospitalization	A1304	Telemetry Discrepancy	19/10/2023	19/10/2023
							F1905	Device Revision or Replacement	A1304	Telemetry Discrepancy	19/10/2023	19/10/2023
14426	2023		47265	Dual-chamber implantable pacemaker, rate-responsive	CT141	Electrodes/Leads and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	07/11/2023	07/11/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	07/11/2023	07/11/2023
					CT145	Cardiovascular devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	07/11/2023	07/11/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	07/11/2023	07/11/2023
					CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	07/11/2023	07/11/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	07/11/2023	07/11/2023
					CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	07/11/2023	07/11/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	07/11/2023	07/11/2023
					CT233	Surgical	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	07/11/2023	07/11/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	07/11/2023	07/11/2023
					CT2406	Implantable	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	07/11/2023	07/11/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	07/11/2023	07/11/2023
					CT2407	Active implantable	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	07/11/2023	07/11/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	07/11/2023	07/11/2023
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	07/11/2023	07/11/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	07/11/2023	07/11/2023
					CT2727	Active implant devices/ancillaries	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	07/11/2023	07/11/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	07/11/2023	07/11/2023
					CT2750	Implantable cardiac electrical stimulators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	07/11/2023	07/11/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	07/11/2023	07/11/2023
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	07/11/2023	07/11/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	07/11/2023	07/11/2023
					CT2985	Active	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	07/11/2023	07/11/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	07/11/2023	07/11/2023
					CT3001	Implantable metal	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	07/11/2023	07/11/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	07/11/2023	07/11/2023
					CT3002	Implantable synthetic polymer	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	07/11/2023	07/11/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	07/11/2023	07/11/2023
					CT3006	Implantable inorganic material	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	07/11/2023	07/11/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	07/11/2023	07/11/2023
					CT314	Electrically powered	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	07/11/2023	07/11/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	07/11/2023	07/11/2023
					CT321	Long-term surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	07/11/2023	07/11/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	07/11/2023	07/11/2023
					CT449	Generators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	07/11/2023	07/11/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	07/11/2023	07/11/2023
					CT981	Single use	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	07/11/2023	07/11/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	07/11/2023	07/11/2023
					CT983	Surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	07/11/2023	07/11/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	07/11/2023	07/11/2023
14427	2023	All other reportable incidents	35852	Single-chamber implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A070910	Undersensing		
									A0712	Pacing Problem		
							F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A070910	Undersensing		
									A0712	Pacing Problem		
14428	2023	All other reportable incidents	35223	Endocardial/interventricular septal pacing lead	CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070101	Failure to Capture		
									A150207	Difficult to Remove		
							F1901	Additional Surgery	A070101	Failure to Capture		
									A150207	Difficult to Remove		
14429	2023		47265	Dual-chamber implantable pacemaker, rate-responsive	CT1685	Cardiac pacemakers	E0112	Dizziness	A0203	Defective Device	25/09/2023	25/09/2023
									A070909	Oversensing	25/09/2023	25/09/2023
									A071204	No Pacing	25/09/2023	25/09/2023
							E011902	Presyncope	A0203	Defective Device	25/09/2023	25/09/2023
									A070909	Oversensing	25/09/2023	25/09/2023
									A071204	No Pacing	25/09/2023	25/09/2023
							F1901	Additional Surgery	A0203	Defective Device	25/09/2023	25/09/2023
									A070909	Oversensing	25/09/2023	25/09/2023
									A071204	No Pacing	25/09/2023	25/09/2023
							F1903	Device Explantation	A0203	Defective Device	25/09/2023	25/09/2023
									A070909	Oversensing	25/09/2023	25/09/2023
									A071204	No Pacing	25/09/2023	25/09/2023
							F1905	Device Revision or Replacement	A0203	Defective Device	25/09/2023	25/09/2023
									A070909	Oversensing	25/09/2023	25/09/2023
									A071204	No Pacing	25/09/2023	25/09/2023
					CT257	Implantable cardiac pacemakers	E0112	Dizziness	A0203	Defective Device	25/09/2023	25/09/2023
									A070909	Oversensing	25/09/2023	25/09/2023
									A071204	No Pacing	25/09/2023	25/09/2023
							E011902	Presyncope	A0203	Defective Device	25/09/2023	25/09/2023
									A070909	Oversensing	25/09/2023	25/09/2023
									A071204	No Pacing	25/09/2023	25/09/2023
							F1901	Additional Surgery	A0203	Defective Device	25/09/2023	25/09/2023
									A070909	Oversensing	25/09/2023	25/09/2023
									A071204	No Pacing	25/09/2023	25/09/2023
							F1903	Device Explantation	A0203	Defective Device	25/09/2023	25/09/2023
									A070909	Oversensing	25/09/2023	25/09/2023
									A071204	No Pacing	25/09/2023	25/09/2023
							F1905	Device Revision or Replacement	A0203	Defective Device	25/09/2023	25/09/2023
									A070909	Oversensing	25/09/2023	25/09/2023
									A071204	No Pacing	25/09/2023	25/09/2023

					CT2866	Implantable cardiac electrical stimulators	E0112	Dizziness	A0203	Defective Device	25/09/2023	25/09/2023
									A070909	Oversensing	25/09/2023	25/09/2023
									A071204	No Pacing	25/09/2023	25/09/2023
							E011902	Presyncope	A0203	Defective Device	25/09/2023	25/09/2023
									A070909	Oversensing	25/09/2023	25/09/2023
									A071204	No Pacing	25/09/2023	25/09/2023
							F1901	Additional Surgery	A0203	Defective Device	25/09/2023	25/09/2023
									A070909	Oversensing	25/09/2023	25/09/2023
									A071204	No Pacing	25/09/2023	25/09/2023
							F1903	Device Explantation	A0203	Defective Device	25/09/2023	25/09/2023
									A070909	Oversensing	25/09/2023	25/09/2023
									A071204	No Pacing	25/09/2023	25/09/2023
							F1905	Device Revision or Replacement	A0203	Defective Device	25/09/2023	25/09/2023
									A070909	Oversensing	25/09/2023	25/09/2023
									A071204	No Pacing	25/09/2023	25/09/2023
14430	2023	All other reportable incidents	47265	Dual-chamber implantable pacemaker, rate-responsive	CT141	Electrodes/Leads and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
									A071102	Failure to Interrogate		
									A071204	No Pacing		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
									A071102	Failure to Interrogate		
									A071204	No Pacing		
					CT145	Cardiovascular devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
									A071102	Failure to Interrogate		
									A071204	No Pacing		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
									A071102	Failure to Interrogate		
									A071204	No Pacing		
					CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
									A071102	Failure to Interrogate		
									A071204	No Pacing		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
									A071102	Failure to Interrogate		
									A071204	No Pacing		
					CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
									A071102	Failure to Interrogate		
									A071204	No Pacing		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
									A071102	Failure to Interrogate		
									A071204	No Pacing		
					CT233	Surgical	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
									A071102	Failure to Interrogate		
									A071204	No Pacing		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
									A071102	Failure to Interrogate		
									A071204	No Pacing		
					CT2406	Implantable	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
									A071102	Failure to Interrogate		
									A071204	No Pacing		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
									A071102	Failure to Interrogate		
									A071204	No Pacing		
					CT2407	Active implantable	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
									A071102	Failure to Interrogate		
									A071204	No Pacing		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
									A071102	Failure to Interrogate		
									A071204	No Pacing		
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
									A071102	Failure to Interrogate		
									A071204	No Pacing		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
									A071102	Failure to Interrogate		
									A071204	No Pacing		
					CT2727	Active implant devices/ancillaries	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
									A071102	Failure to Interrogate		
									A071204	No Pacing		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
									A071102	Failure to Interrogate		
									A071204	No Pacing		
					CT2750	Implantable cardiac electrical stimulators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
									A071102	Failure to Interrogate		
									A071204	No Pacing		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
									A071102	Failure to Interrogate		
									A071204	No Pacing		
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
									A071102	Failure to Interrogate		
									A071204	No Pacing		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
									A071102	Failure to Interrogate		
									A071204	No Pacing		
					CT2985	Active	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
									A071102	Failure to Interrogate		
									A071204	No Pacing		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
									A071102	Failure to Interrogate		
									A071204	No Pacing		
					CT3001	Implantable metal	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		

									A071102	Failure to Interrogate		
									A071204	No Pacing		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
									A071102	Failure to Interrogate		
									A071204	No Pacing		
					CT3002	Implantable synthetic polymer	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
									A071102	Failure to Interrogate		
									A071204	No Pacing		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
									A071102	Failure to Interrogate		
									A071204	No Pacing		
					CT3006	Implantable inorganic material	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
									A071102	Failure to Interrogate		
									A071204	No Pacing		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
									A071102	Failure to Interrogate		
									A071204	No Pacing		
					CT314	Electrically powered	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
									A071102	Failure to Interrogate		
									A071204	No Pacing		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
									A071102	Failure to Interrogate		
									A071204	No Pacing		
					CT321	Long-term surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
									A071102	Failure to Interrogate		
									A071204	No Pacing		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
									A071102	Failure to Interrogate		
									A071204	No Pacing		
					CT449	Generators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
									A071102	Failure to Interrogate		
									A071204	No Pacing		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
									A071102	Failure to Interrogate		
									A071204	No Pacing		
					CT981	Single use	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
									A071102	Failure to Interrogate		
									A071204	No Pacing		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
									A071102	Failure to Interrogate		
									A071204	No Pacing		
					CT983	Surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
									A071102	Failure to Interrogate		
									A071204	No Pacing		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
									A071102	Failure to Interrogate		
									A071204	No Pacing		
14431	2023	All other reportable incidents	47265	Dual-chamber implantable pacemaker, rate-responsive	CT141	Electrodes/Leads and associated devices	E060106	Heart Block	A070101	Failure to Capture	23/02/2023	23/02/2023
									A070102	High Capture Threshold	23/02/2023	23/02/2023
									A070901	Decreased Sensitivity	23/02/2023	23/02/2023
									A072201	High Impedance	23/02/2023	23/02/2023
							F08	Hospitalization or Prolonged Hospitalization	A070101	Failure to Capture	23/02/2023	23/02/2023
									A070102	High Capture Threshold	23/02/2023	23/02/2023
									A070901	Decreased Sensitivity	23/02/2023	23/02/2023
									A072201	High Impedance	23/02/2023	23/02/2023
							F1905	Device Revision or Replacement	A070101	Failure to Capture	23/02/2023	23/02/2023
									A070102	High Capture Threshold	23/02/2023	23/02/2023
									A070901	Decreased Sensitivity	23/02/2023	23/02/2023
									A072201	High Impedance	23/02/2023	23/02/2023
					CT145	Cardiovascular devices	E060106	Heart Block	A070101	Failure to Capture	23/02/2023	23/02/2023
									A070102	High Capture Threshold	23/02/2023	23/02/2023
									A070901	Decreased Sensitivity	23/02/2023	23/02/2023
									A072201	High Impedance	23/02/2023	23/02/2023
							F08	Hospitalization or Prolonged Hospitalization	A070101	Failure to Capture	23/02/2023	23/02/2023
									A070102	High Capture Threshold	23/02/2023	23/02/2023
									A070901	Decreased Sensitivity	23/02/2023	23/02/2023
									A072201	High Impedance	23/02/2023	23/02/2023
							F1905	Device Revision or Replacement	A070101	Failure to Capture	23/02/2023	23/02/2023
									A070102	High Capture Threshold	23/02/2023	23/02/2023
									A070901	Decreased Sensitivity	23/02/2023	23/02/2023
									A072201	High Impedance	23/02/2023	23/02/2023
					CT1684	Cardiac pacemakers and associated devices	E060106	Heart Block	A070101	Failure to Capture	23/02/2023	23/02/2023
									A070102	High Capture Threshold	23/02/2023	23/02/2023
									A070901	Decreased Sensitivity	23/02/2023	23/02/2023
									A072201	High Impedance	23/02/2023	23/02/2023
							F08	Hospitalization or Prolonged Hospitalization	A070101	Failure to Capture	23/02/2023	23/02/2023
									A070102	High Capture Threshold	23/02/2023	23/02/2023
									A070901	Decreased Sensitivity	23/02/2023	23/02/2023
									A072201	High Impedance	23/02/2023	23/02/2023
							F1905	Device Revision or Replacement	A070101	Failure to Capture	23/02/2023	23/02/2023
									A070102	High Capture Threshold	23/02/2023	23/02/2023
									A070901	Decreased Sensitivity	23/02/2023	23/02/2023
									A072201	High Impedance	23/02/2023	23/02/2023
					CT1685	Cardiac pacemakers	E060106	Heart Block	A070101	Failure to Capture	23/02/2023	23/02/2023
									A070102	High Capture Threshold	23/02/2023	23/02/2023
									A070901	Decreased Sensitivity	23/02/2023	23/02/2023
									A072201	High Impedance	23/02/2023	23/02/2023
							F08	Hospitalization or Prolonged Hospitalization	A070101	Failure to Capture	23/02/2023	23/02/2023

									A070102	High Capture Threshold	23/02/2023	23/02/2023
									A070901	Decreased Sensitivity	23/02/2023	23/02/2023
									A072201	High Impedance	23/02/2023	23/02/2023
								F1905	Device Revision or Replacement	A070101	Failure to Capture	23/02/2023	23/02/2023
										A070102	High Capture Threshold	23/02/2023	23/02/2023
										A070901	Decreased Sensitivity	23/02/2023	23/02/2023
										A072201	High Impedance	23/02/2023	23/02/2023
					CT233	Surgical	E060106	Heart Block		A070101	Failure to Capture	23/02/2023	23/02/2023
										A070102	High Capture Threshold	23/02/2023	23/02/2023
										A070901	Decreased Sensitivity	23/02/2023	23/02/2023
										A072201	High Impedance	23/02/2023	23/02/2023
								F08	Hospitalization or Prolonged Hospitalization	A070101	Failure to Capture	23/02/2023	23/02/2023
										A070102	High Capture Threshold	23/02/2023	23/02/2023
										A070901	Decreased Sensitivity	23/02/2023	23/02/2023
										A072201	High Impedance	23/02/2023	23/02/2023
								F1905	Device Revision or Replacement	A070101	Failure to Capture	23/02/2023	23/02/2023
										A070102	High Capture Threshold	23/02/2023	23/02/2023
										A070901	Decreased Sensitivity	23/02/2023	23/02/2023
										A072201	High Impedance	23/02/2023	23/02/2023
					CT2406	Implantable	E060106	Heart Block		A070101	Failure to Capture	23/02/2023	23/02/2023
										A070102	High Capture Threshold	23/02/2023	23/02/2023
										A070901	Decreased Sensitivity	23/02/2023	23/02/2023
										A072201	High Impedance	23/02/2023	23/02/2023
								F08	Hospitalization or Prolonged Hospitalization	A070101	Failure to Capture	23/02/2023	23/02/2023
										A070102	High Capture Threshold	23/02/2023	23/02/2023
										A070901	Decreased Sensitivity	23/02/2023	23/02/2023
										A072201	High Impedance	23/02/2023	23/02/2023
								F1905	Device Revision or Replacement	A070101	Failure to Capture	23/02/2023	23/02/2023
										A070102	High Capture Threshold	23/02/2023	23/02/2023
										A070901	Decreased Sensitivity	23/02/2023	23/02/2023
										A072201	High Impedance	23/02/2023	23/02/2023
					CT2407	Active implantable	E060106	Heart Block		A070101	Failure to Capture	23/02/2023	23/02/2023
										A070102	High Capture Threshold	23/02/2023	23/02/2023
										A070901	Decreased Sensitivity	23/02/2023	23/02/2023
										A072201	High Impedance	23/02/2023	23/02/2023
								F08	Hospitalization or Prolonged Hospitalization	A070101	Failure to Capture	23/02/2023	23/02/2023
										A070102	High Capture Threshold	23/02/2023	23/02/2023
										A070901	Decreased Sensitivity	23/02/2023	23/02/2023
										A072201	High Impedance	23/02/2023	23/02/2023
								F1905	Device Revision or Replacement	A070101	Failure to Capture	23/02/2023	23/02/2023
										A070102	High Capture Threshold	23/02/2023	23/02/2023
										A070901	Decreased Sensitivity	23/02/2023	23/02/2023
										A072201	High Impedance	23/02/2023	23/02/2023
					CT257	Implantable cardiac pacemakers	E060106	Heart Block		A070101	Failure to Capture	23/02/2023	23/02/2023
										A070102	High Capture Threshold	23/02/2023	23/02/2023
										A070901	Decreased Sensitivity	23/02/2023	23/02/2023
										A072201	High Impedance	23/02/2023	23/02/2023
								F08	Hospitalization or Prolonged Hospitalization	A070101	Failure to Capture	23/02/2023	23/02/2023
										A070102	High Capture Threshold	23/02/2023	23/02/2023
										A070901	Decreased Sensitivity	23/02/2023	23/02/2023
										A072201	High Impedance	23/02/2023	23/02/2023
								F1905	Device Revision or Replacement	A070101	Failure to Capture	23/02/2023	23/02/2023
										A070102	High Capture Threshold	23/02/2023	23/02/2023
										A070901	Decreased Sensitivity	23/02/2023	23/02/2023
										A072201	High Impedance	23/02/2023	23/02/2023
					CT2727	Active implant devices/ancillaries	E060106	Heart Block		A070101	Failure to Capture	23/02/2023	23/02/2023
										A070102	High Capture Threshold	23/02/2023	23/02/2023
										A070901	Decreased Sensitivity	23/02/2023	23/02/2023
										A072201	High Impedance	23/02/2023	23/02/2023
								F08	Hospitalization or Prolonged Hospitalization	A070101	Failure to Capture	23/02/2023	23/02/2023
										A070102	High Capture Threshold	23/02/2023	23/02/2023
										A070901	Decreased Sensitivity	23/02/2023	23/02/2023
										A072201	High Impedance	23/02/2023	23/02/2023
								F1905	Device Revision or Replacement	A070101	Failure to Capture	23/02/2023	23/02/2023
										A070102	High Capture Threshold	23/02/2023	23/02/2023
										A070901	Decreased Sensitivity	23/02/2023	23/02/2023
										A072201	High Impedance	23/02/2023	23/02/2023
					CT2750	Implantable cardiac electrical stimulators and associated devices	E060106	Heart Block		A070101	Failure to Capture	23/02/2023	23/02/2023
										A070102	High Capture Threshold	23/02/2023	23/02/2023
										A070901	Decreased Sensitivity	23/02/2023	23/02/2023
										A072201	High Impedance	23/02/2023	23/02/2023
								F08	Hospitalization or Prolonged Hospitalization	A070101	Failure to Capture	23/02/2023	23/02/2023
										A070102	High Capture Threshold	23/02/2023	23/02/2023
										A070901	Decreased Sensitivity	23/02/2023	23/02/2023
										A072201	High Impedance	23/02/2023	23/02/2023
								F1905	Device Revision or Replacement	A070101	Failure to Capture	23/02/2023	23/02/2023
										A070102	High Capture Threshold	23/02/2023	23/02/2023
										A070901	Decreased Sensitivity	23/02/2023	23/02/2023
										A072201	High Impedance	23/02/2023	23/02/2023
					CT2866	Implantable cardiac electrical stimulators	E060106	Heart Block		A070101	Failure to Capture	23/02/2023	23/02/2023
										A070102	High Capture Threshold	23/02/2023	23/02/2023
										A070901	Decreased Sensitivity	23/02/2023	23/02/2023
										A072201	High Impedance	23/02/2023	23/02/2023
								F08	Hospitalization or Prolonged Hospitalization	A070101	Failure to Capture	23/02/2023	23/02/2023
										A070102	High Capture Threshold	23/02/2023	23/02/2023
										A070901	Decreased Sensitivity	23/02/2023	23/02/2023
										A072201	High Impedance	23/02/2023	23/02/2023
								F1905	Device Revision or Replacement	A070101	Failure to Capture	23/02/2023	23/02/2023

								A070102	High Capture Threshold	23/02/2023	23/02/2023
								A070901	Decreased Sensitivity	23/02/2023	23/02/2023
								A072201	High Impedance	23/02/2023	23/02/2023
					CT2985	Active	E060106	Heart Block	A070101	Failure to Capture	23/02/2023	23/02/2023
									A070102	High Capture Threshold	23/02/2023	23/02/2023
									A070901	Decreased Sensitivity	23/02/2023	23/02/2023
									A072201	High Impedance	23/02/2023	23/02/2023
							F08	Hospitalization or Prolonged Hospitalization	A070101	Failure to Capture	23/02/2023	23/02/2023
									A070102	High Capture Threshold	23/02/2023	23/02/2023
									A070901	Decreased Sensitivity	23/02/2023	23/02/2023
									A072201	High Impedance	23/02/2023	23/02/2023
							F1905	Device Revision or Replacement	A070101	Failure to Capture	23/02/2023	23/02/2023
									A070102	High Capture Threshold	23/02/2023	23/02/2023
									A070901	Decreased Sensitivity	23/02/2023	23/02/2023
									A072201	High Impedance	23/02/2023	23/02/2023
					CT3001	Implantable metal	E060106	Heart Block	A070101	Failure to Capture	23/02/2023	23/02/2023
									A070102	High Capture Threshold	23/02/2023	23/02/2023
									A070901	Decreased Sensitivity	23/02/2023	23/02/2023
									A072201	High Impedance	23/02/2023	23/02/2023
							F08	Hospitalization or Prolonged Hospitalization	A070101	Failure to Capture	23/02/2023	23/02/2023
									A070102	High Capture Threshold	23/02/2023	23/02/2023
									A070901	Decreased Sensitivity	23/02/2023	23/02/2023
									A072201	High Impedance	23/02/2023	23/02/2023
							F1905	Device Revision or Replacement	A070101	Failure to Capture	23/02/2023	23/02/2023
									A070102	High Capture Threshold	23/02/2023	23/02/2023
									A070901	Decreased Sensitivity	23/02/2023	23/02/2023
									A072201	High Impedance	23/02/2023	23/02/2023
					CT3002	Implantable synthetic polymer	E060106	Heart Block	A070101	Failure to Capture	23/02/2023	23/02/2023
									A070102	High Capture Threshold	23/02/2023	23/02/2023
									A070901	Decreased Sensitivity	23/02/2023	23/02/2023
									A072201	High Impedance	23/02/2023	23/02/2023
							F08	Hospitalization or Prolonged Hospitalization	A070101	Failure to Capture	23/02/2023	23/02/2023
									A070102	High Capture Threshold	23/02/2023	23/02/2023
									A070901	Decreased Sensitivity	23/02/2023	23/02/2023
									A072201	High Impedance	23/02/2023	23/02/2023
							F1905	Device Revision or Replacement	A070101	Failure to Capture	23/02/2023	23/02/2023
									A070102	High Capture Threshold	23/02/2023	23/02/2023
									A070901	Decreased Sensitivity	23/02/2023	23/02/2023
									A072201	High Impedance	23/02/2023	23/02/2023
					CT3006	Implantable inorganic material	E060106	Heart Block	A070101	Failure to Capture	23/02/2023	23/02/2023
									A070102	High Capture Threshold	23/02/2023	23/02/2023
									A070901	Decreased Sensitivity	23/02/2023	23/02/2023
									A072201	High Impedance	23/02/2023	23/02/2023
							F08	Hospitalization or Prolonged Hospitalization	A070101	Failure to Capture	23/02/2023	23/02/2023
									A070102	High Capture Threshold	23/02/2023	23/02/2023
									A070901	Decreased Sensitivity	23/02/2023	23/02/2023
									A072201	High Impedance	23/02/2023	23/02/2023
							F1905	Device Revision or Replacement	A070101	Failure to Capture	23/02/2023	23/02/2023
									A070102	High Capture Threshold	23/02/2023	23/02/2023
									A070901	Decreased Sensitivity	23/02/2023	23/02/2023
									A072201	High Impedance	23/02/2023	23/02/2023
					CT314	Electrically powered	E060106	Heart Block	A070101	Failure to Capture	23/02/2023	23/02/2023
									A070102	High Capture Threshold	23/02/2023	23/02/2023
									A070901	Decreased Sensitivity	23/02/2023	23/02/2023
									A072201	High Impedance	23/02/2023	23/02/2023
							F08	Hospitalization or Prolonged Hospitalization	A070101	Failure to Capture	23/02/2023	23/02/2023
									A070102	High Capture Threshold	23/02/2023	23/02/2023
									A070901	Decreased Sensitivity	23/02/2023	23/02/2023
									A072201	High Impedance	23/02/2023	23/02/2023
							F1905	Device Revision or Replacement	A070101	Failure to Capture	23/02/2023	23/02/2023
									A070102	High Capture Threshold	23/02/2023	23/02/2023
									A070901	Decreased Sensitivity	23/02/2023	23/02/2023
									A072201	High Impedance	23/02/2023	23/02/2023
					CT321	Long-term surgical invasive	E060106	Heart Block	A070101	Failure to Capture	23/02/2023	23/02/2023
									A070102	High Capture Threshold	23/02/2023	23/02/2023
									A070901	Decreased Sensitivity	23/02/2023	23/02/2023
									A072201	High Impedance	23/02/2023	23/02/2023
							F08	Hospitalization or Prolonged Hospitalization	A070101	Failure to Capture	23/02/2023	23/02/2023
									A070102	High Capture Threshold	23/02/2023	23/02/2023
									A070901	Decreased Sensitivity	23/02/2023	23/02/2023
									A072201	High Impedance	23/02/2023	23/02/2023
							F1905	Device Revision or Replacement	A070101	Failure to Capture	23/02/2023	23/02/2023
									A070102	High Capture Threshold	23/02/2023	23/02/2023
									A070901	Decreased Sensitivity	23/02/2023	23/02/2023
									A072201	High Impedance	23/02/2023	23/02/2023
					CT449	Generators and associated devices	E060106	Heart Block	A070101	Failure to Capture	23/02/2023	23/02/2023
									A070102	High Capture Threshold	23/02/2023	23/02/2023
									A070901	Decreased Sensitivity	23/02/2023	23/02/2023
									A072201	High Impedance	23/02/2023	23/02/2023
							F08	Hospitalization or Prolonged Hospitalization	A070101	Failure to Capture	23/02/2023	23/02/2023
									A070102	High Capture Threshold	23/02/2023	23/02/2023
									A070901	Decreased Sensitivity	23/02/2023	23/02/2023
									A072201	High Impedance	23/02/2023	23/02/2023
							F1905	Device Revision or Replacement	A070101	Failure to Capture	23/02/2023	23/02/2023
									A070102	High Capture Threshold	23/02/2023	23/02/2023
									A070901	Decreased Sensitivity	23/02/2023	23/02/2023
									A072201	High Impedance	23/02/2023	23/02/2023
					CT981	Single use	E060106	Heart Block	A070101	Failure to Capture	23/02/2023	23/02/2023

								A070102	High Capture Threshold	23/02/2023	23/02/2023
								A070901	Decreased Sensitivity	23/02/2023	23/02/2023
								A072201	High Impedance	23/02/2023	23/02/2023
						F08	Hospitalization or Prolonged Hospitalization	A070101	Failure to Capture	23/02/2023	23/02/2023
								A070102	High Capture Threshold	23/02/2023	23/02/2023
								A070901	Decreased Sensitivity	23/02/2023	23/02/2023
								A072201	High Impedance	23/02/2023	23/02/2023
						F1905	Device Revision or Replacement	A070101	Failure to Capture	23/02/2023	23/02/2023
								A070102	High Capture Threshold	23/02/2023	23/02/2023
								A070901	Decreased Sensitivity	23/02/2023	23/02/2023
								A072201	High Impedance	23/02/2023	23/02/2023
				CT983	Surgical invasive	E060106	Heart Block	A070101	Failure to Capture	23/02/2023	23/02/2023
								A070102	High Capture Threshold	23/02/2023	23/02/2023
								A070901	Decreased Sensitivity	23/02/2023	23/02/2023
								A072201	High Impedance	23/02/2023	23/02/2023
						F08	Hospitalization or Prolonged Hospitalization	A070101	Failure to Capture	23/02/2023	23/02/2023
								A070102	High Capture Threshold	23/02/2023	23/02/2023
								A070901	Decreased Sensitivity	23/02/2023	23/02/2023
								A072201	High Impedance	23/02/2023	23/02/2023
						F1905	Device Revision or Replacement	A070101	Failure to Capture	23/02/2023	23/02/2023
								A070102	High Capture Threshold	23/02/2023	23/02/2023
								A070901	Decreased Sensitivity	23/02/2023	23/02/2023
								A072201	High Impedance	23/02/2023	23/02/2023
14432	2023	All other reportable incidents	37265	Dual-chamber implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E0119	Loss of Consciousness	A071201	Failure to Convert Rhythm		
									A071302	Inappropriate/Inadequate Shock/Stimulation		
							E060102	Atrial Fibrillation	A071201	Failure to Convert Rhythm		
									A071302	Inappropriate/Inadequate Shock/Stimulation		
							E2104	Electric Shock	A071201	Failure to Convert Rhythm		
									A071302	Inappropriate/Inadequate Shock/Stimulation		
							F08	Hospitalization or Prolonged Hospitalization	A071201	Failure to Convert Rhythm		
									A071302	Inappropriate/Inadequate Shock/Stimulation		
							F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A071201	Failure to Convert Rhythm		
									A071302	Inappropriate/Inadequate Shock/Stimulation		
							F1901	Additional Surgery	A071201	Failure to Convert Rhythm		
									A071302	Inappropriate/Inadequate Shock/Stimulation		
							F2301	Additional Device Required	A071201	Failure to Convert Rhythm		
									A071302	Inappropriate/Inadequate Shock/Stimulation		
14433	2023		47265	Dual-chamber implantable pacemaker, rate-responsive	CT141	Electrodes/Leads and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A0203	Defective Device		
									A11	Computer Software Problem		
							F08	Hospitalization or Prolonged Hospitalization	A0203	Defective Device		
									A11	Computer Software Problem		
					CT145	Cardiovascular devices	E2403	No Clinical Signs, Symptoms or Conditions	A0203	Defective Device		
									A11	Computer Software Problem		
							F08	Hospitalization or Prolonged Hospitalization	A0203	Defective Device		
									A11	Computer Software Problem		
					CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A0203	Defective Device		
									A11	Computer Software Problem		
							F08	Hospitalization or Prolonged Hospitalization	A0203	Defective Device		
									A11	Computer Software Problem		
					CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A0203	Defective Device		
									A11	Computer Software Problem		
							F08	Hospitalization or Prolonged Hospitalization	A0203	Defective Device		
									A11	Computer Software Problem		
					CT233	Surgical	E2403	No Clinical Signs, Symptoms or Conditions	A0203	Defective Device		
									A11	Computer Software Problem		
							F08	Hospitalization or Prolonged Hospitalization	A0203	Defective Device		
									A11	Computer Software Problem		
					CT2406	Implantable	E2403	No Clinical Signs, Symptoms or Conditions	A0203	Defective Device		
									A11	Computer Software Problem		
							F08	Hospitalization or Prolonged Hospitalization	A0203	Defective Device		
									A11	Computer Software Problem		
					CT2407	Active implantable	E2403	No Clinical Signs, Symptoms or Conditions	A0203	Defective Device		
									A11	Computer Software Problem		
							F08	Hospitalization or Prolonged Hospitalization	A0203	Defective Device		
									A11	Computer Software Problem		
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A0203	Defective Device		
									A11	Computer Software Problem		
							F08	Hospitalization or Prolonged Hospitalization	A0203	Defective Device		
									A11	Computer Software Problem		
					CT2727	Active implant devices/ancillaries	E2403	No Clinical Signs, Symptoms or Conditions	A0203	Defective Device		
									A11	Computer Software Problem		
							F08	Hospitalization or Prolonged Hospitalization	A0203	Defective Device		
									A11	Computer Software Problem		
					CT2750	Implantable cardiac electrical stimulators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A0203	Defective Device		
									A11	Computer Software Problem		
							F08	Hospitalization or Prolonged Hospitalization	A0203	Defective Device		
									A11	Computer Software Problem		
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A0203	Defective Device		
									A11	Computer Software Problem		
							F08	Hospitalization or Prolonged Hospitalization	A0203	Defective Device		
									A11	Computer Software Problem		
					CT2985	Active	E2403	No Clinical Signs, Symptoms or Conditions	A0203	Defective Device		
									A11	Computer Software Problem		
							F08	Hospitalization or Prolonged Hospitalization	A0203	Defective Device		
									A11	Computer Software Problem		
					CT3001	Implantable metal	E2403	No Clinical Signs, Symptoms or Conditions	A0203	Defective Device		
									A11	Computer Software Problem		
							F08	Hospitalization or Prolonged Hospitalization	A0203	Defective Device		

									A11	Computer Software Problem		
					CT3002	Implantable synthetic polymer	E2403	No Clinical Signs, Symptoms or Conditions	A0203	Defective Device		
									A11	Computer Software Problem		
							F08	Hospitalization or Prolonged Hospitalization	A0203	Defective Device		
									A11	Computer Software Problem		
					CT3006	Implantable inorganic material	E2403	No Clinical Signs, Symptoms or Conditions	A0203	Defective Device		
									A11	Computer Software Problem		
							F08	Hospitalization or Prolonged Hospitalization	A0203	Defective Device		
									A11	Computer Software Problem		
					CT314	Electrically powered	E2403	No Clinical Signs, Symptoms or Conditions	A0203	Defective Device		
									A11	Computer Software Problem		
							F08	Hospitalization or Prolonged Hospitalization	A0203	Defective Device		
									A11	Computer Software Problem		
					CT321	Long-term surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A0203	Defective Device		
									A11	Computer Software Problem		
							F08	Hospitalization or Prolonged Hospitalization	A0203	Defective Device		
									A11	Computer Software Problem		
					CT449	Generators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A0203	Defective Device		
									A11	Computer Software Problem		
							F08	Hospitalization or Prolonged Hospitalization	A0203	Defective Device		
									A11	Computer Software Problem		
					CT981	Single use	E2403	No Clinical Signs, Symptoms or Conditions	A0203	Defective Device		
									A11	Computer Software Problem		
							F08	Hospitalization or Prolonged Hospitalization	A0203	Defective Device		
									A11	Computer Software Problem		
					CT983	Surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A0203	Defective Device		
									A11	Computer Software Problem		
							F08	Hospitalization or Prolonged Hospitalization	A0203	Defective Device		
									A11	Computer Software Problem		
14434	2023	All other reportable incidents	35223	Endocardial/interventricular septal pacing lead	CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070909	Oversensing		
									A090801	Signal Artifact/Noise		
							F12	Serious Injury/ Illness/ Impairment	A070909	Oversensing		
									A090801	Signal Artifact/Noise		
							F23	Unexpected Medical Intervention	A070909	Oversensing		
									A090801	Signal Artifact/Noise		
14435	2023	All other reportable incidents	47265	Dual-chamber implantable pacemaker, rate-responsive	CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
14436	2023	All other reportable incidents	35852	Single-chamber implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	25/01/2024	25/01/2024
									A0908	Incorrect, Inadequate or Imprecise Result or Readings	25/01/2024	25/01/2024
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	25/01/2024	25/01/2024
									A0908	Incorrect, Inadequate or Imprecise Result or Readings	25/01/2024	25/01/2024
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	25/01/2024	25/01/2024
									A0908	Incorrect, Inadequate or Imprecise Result or Readings	25/01/2024	25/01/2024
14437	2023	All other reportable incidents	47270	Cardiac resynchronization therapy implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A070909	Oversensing		
									A090801	Signal Artifact/Noise		
							F26	No Health Consequences or Impact	A070909	Oversensing		
									A090801	Signal Artifact/Noise		
14438	2023	All other reportable incidents	47265	Dual-chamber implantable pacemaker, rate-responsive	CT141	Electrodes/Leads and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A072201	High Impedance		
									A1304	Telemetry Discrepancy		
							F26	No Health Consequences or Impact	A072201	High Impedance		
									A1304	Telemetry Discrepancy		
					CT145	Cardiovascular devices	E2403	No Clinical Signs, Symptoms or Conditions	A072201	High Impedance		
									A1304	Telemetry Discrepancy		
							F26	No Health Consequences or Impact	A072201	High Impedance		
									A1304	Telemetry Discrepancy		
					CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A072201	High Impedance		
									A1304	Telemetry Discrepancy		
							F26	No Health Consequences or Impact	A072201	High Impedance		
									A1304	Telemetry Discrepancy		
					CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A072201	High Impedance		
									A1304	Telemetry Discrepancy		
							F26	No Health Consequences or Impact	A072201	High Impedance		
									A1304	Telemetry Discrepancy		
					CT233	Surgical	E2403	No Clinical Signs, Symptoms or Conditions	A072201	High Impedance		
									A1304	Telemetry Discrepancy		
							F26	No Health Consequences or Impact	A072201	High Impedance		
									A1304	Telemetry Discrepancy		
					CT2406	Implantable	E2403	No Clinical Signs, Symptoms or Conditions	A072201	High Impedance		
									A1304	Telemetry Discrepancy		
							F26	No Health Consequences or Impact	A072201	High Impedance		
									A1304	Telemetry Discrepancy		
					CT2407	Active implantable	E2403	No Clinical Signs, Symptoms or Conditions	A072201	High Impedance		
									A1304	Telemetry Discrepancy		
							F26	No Health Consequences or Impact	A072201	High Impedance		
									A1304	Telemetry Discrepancy		
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A072201	High Impedance		
									A1304	Telemetry Discrepancy		
							F26	No Health Consequences or Impact	A072201	High Impedance		
									A1304	Telemetry Discrepancy		
					CT2727	Active implant devices/ancillaries	E2403	No Clinical Signs, Symptoms or Conditions	A072201	High Impedance		
									A1304	Telemetry Discrepancy		

							F26	No Health Consequences or Impact	A072201	High Impedance		
									A1304	Telemetry Discrepancy		
					CT2750	Implantable cardiac electrical stimulators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A072201	High Impedance		
									A1304	Telemetry Discrepancy		
							F26	No Health Consequences or Impact	A072201	High Impedance		
									A1304	Telemetry Discrepancy		
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A072201	High Impedance		
									A1304	Telemetry Discrepancy		
							F26	No Health Consequences or Impact	A072201	High Impedance		
									A1304	Telemetry Discrepancy		
					CT2985	Active	E2403	No Clinical Signs, Symptoms or Conditions	A072201	High Impedance		
									A1304	Telemetry Discrepancy		
							F26	No Health Consequences or Impact	A072201	High Impedance		
									A1304	Telemetry Discrepancy		
					CT3001	Implantable metal	E2403	No Clinical Signs, Symptoms or Conditions	A072201	High Impedance		
									A1304	Telemetry Discrepancy		
							F26	No Health Consequences or Impact	A072201	High Impedance		
									A1304	Telemetry Discrepancy		
					CT3002	Implantable synthetic polymer	E2403	No Clinical Signs, Symptoms or Conditions	A072201	High Impedance		
									A1304	Telemetry Discrepancy		
							F26	No Health Consequences or Impact	A072201	High Impedance		
									A1304	Telemetry Discrepancy		
					CT3006	Implantable inorganic material	E2403	No Clinical Signs, Symptoms or Conditions	A072201	High Impedance		
									A1304	Telemetry Discrepancy		
							F26	No Health Consequences or Impact	A072201	High Impedance		
									A1304	Telemetry Discrepancy		
					CT314	Electrically powered	E2403	No Clinical Signs, Symptoms or Conditions	A072201	High Impedance		
									A1304	Telemetry Discrepancy		
							F26	No Health Consequences or Impact	A072201	High Impedance		
									A1304	Telemetry Discrepancy		
					CT321	Long-term surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A072201	High Impedance		
									A1304	Telemetry Discrepancy		
							F26	No Health Consequences or Impact	A072201	High Impedance		
									A1304	Telemetry Discrepancy		
					CT449	Generators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A072201	High Impedance		
									A1304	Telemetry Discrepancy		
							F26	No Health Consequences or Impact	A072201	High Impedance		
									A1304	Telemetry Discrepancy		
					CT981	Single use	E2403	No Clinical Signs, Symptoms or Conditions	A072201	High Impedance		
									A1304	Telemetry Discrepancy		
							F26	No Health Consequences or Impact	A072201	High Impedance		
									A1304	Telemetry Discrepancy		
					CT983	Surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A072201	High Impedance		
									A1304	Telemetry Discrepancy		
							F26	No Health Consequences or Impact	A072201	High Impedance		
									A1304	Telemetry Discrepancy		
14439	2023	All other reportable incidents	47270	Cardiac resynchronization therapy implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E2104	Electric Shock	A070909	Oversensing		
									A071302	Inappropriate/Inadequate Shock/Stimulation		
							F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A070909	Oversensing		
									A071302	Inappropriate/Inadequate Shock/Stimulation		
14440	2023	All other reportable incidents	47265	Dual-chamber implantable pacemaker, rate-responsive	CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
14441	2023		47265	Dual-chamber implantable pacemaker, rate-responsive	CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	10/11/2023	10/11/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	10/11/2023	10/11/2023
							F1903	Device Explantation	A24	No Apparent Device Problem	10/11/2023	10/11/2023
							F1905	Device Revision or Replacement	A24	No Apparent Device Problem	10/11/2023	10/11/2023
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	10/11/2023	10/11/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	10/11/2023	10/11/2023
							F1903	Device Explantation	A24	No Apparent Device Problem	10/11/2023	10/11/2023
							F1905	Device Revision or Replacement	A24	No Apparent Device Problem	10/11/2023	10/11/2023
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	10/11/2023	10/11/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	10/11/2023	10/11/2023
							F1903	Device Explantation	A24	No Apparent Device Problem	10/11/2023	10/11/2023
							F1905	Device Revision or Replacement	A24	No Apparent Device Problem	10/11/2023	10/11/2023
14442	2023	All other reportable incidents	47265	Dual-chamber implantable pacemaker, rate-responsive	CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	25/08/2023	25/08/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	25/08/2023	25/08/2023
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	25/08/2023	25/08/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	25/08/2023	25/08/2023
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	25/08/2023	25/08/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	25/08/2023	25/08/2023
14443	2023		47265	Dual-chamber implantable pacemaker, rate-responsive	CT141	Electrodes/Leads and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	07/11/2023	07/11/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	07/11/2023	07/11/2023
					CT145	Cardiovascular devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	07/11/2023	07/11/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	07/11/2023	07/11/2023
					CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	07/11/2023	07/11/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	07/11/2023	07/11/2023
					CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	07/11/2023	07/11/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	07/11/2023	07/11/2023
					CT233	Surgical	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	07/11/2023	07/11/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	07/11/2023	07/11/2023
					CT2406	Implantable	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	07/11/2023	07/11/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	07/11/2023	07/11/2023
					CT2407	Active implantable	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	07/11/2023	07/11/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	07/11/2023	07/11/2023

					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	07/11/2023	07/11/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	07/11/2023	07/11/2023
					CT2727	Active implant devices/ancillaries	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	07/11/2023	07/11/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	07/11/2023	07/11/2023
					CT2750	Implantable cardiac electrical stimulators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	07/11/2023	07/11/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	07/11/2023	07/11/2023
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	07/11/2023	07/11/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	07/11/2023	07/11/2023
					CT2985	Active	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	07/11/2023	07/11/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	07/11/2023	07/11/2023
					CT3001	Implantable metal	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	07/11/2023	07/11/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	07/11/2023	07/11/2023
					CT3002	Implantable synthetic polymer	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	07/11/2023	07/11/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	07/11/2023	07/11/2023
					CT3006	Implantable inorganic material	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	07/11/2023	07/11/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	07/11/2023	07/11/2023
					CT314	Electrically powered	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	07/11/2023	07/11/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	07/11/2023	07/11/2023
					CT321	Long-term surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	07/11/2023	07/11/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	07/11/2023	07/11/2023
					CT449	Generators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	07/11/2023	07/11/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	07/11/2023	07/11/2023
					CT981	Single use	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	07/11/2023	07/11/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	07/11/2023	07/11/2023
					CT983	Surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	07/11/2023	07/11/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	07/11/2023	07/11/2023
14444	2023	All other reportable incidents	17882	Physiologic-monitoring defibrillation system	CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070803	Failure to Power Up		
							F26	No Health Consequences or Impact	A070803	Failure to Power Up		
14445	2023	All other reportable incidents	35852	Single-chamber implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	12/12/2023	12/12/2023
									A0908	Incorrect, Inadequate or Imprecise Result or Readings	12/12/2023	12/12/2023
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	12/12/2023	12/12/2023
									A0908	Incorrect, Inadequate or Imprecise Result or Readings	12/12/2023	12/12/2023
							F1903	Device Explantation	A070504	Premature Discharge of Battery	12/12/2023	12/12/2023
									A0908	Incorrect, Inadequate or Imprecise Result or Readings	12/12/2023	12/12/2023
14446	2023	All other reportable incidents	47265	Dual-chamber implantable pacemaker, rate-responsive	CT141	Electrodes/Leads and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
					CT145	Cardiovascular devices	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
					CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
					CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
					CT233	Surgical	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
					CT2406	Implantable	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
					CT2407	Active implantable	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
					CT2727	Active implant devices/ancillaries	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
					CT2750	Implantable cardiac electrical stimulators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
					CT2985	Active	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
					CT3001	Implantable metal	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A0705	Battery Problem		

									A070504	Premature Discharge of Battery		
					CT3002	Implantable synthetic polymer	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
					CT3006	Implantable inorganic material	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
					CT314	Electrically powered	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
					CT321	Long-term surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
					CT449	Generators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
					CT981	Single use	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
					CT983	Surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
14447	2023		37265	Dual-chamber implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E0601	Arrhythmia	A071201	Failure to Convert Rhythm		
							F12	Serious Injury/ Illness/ Impairment	A071201	Failure to Convert Rhythm		
14448	2023	All other reportable incidents	47265	Dual-chamber implantable pacemaker, rate-responsive	CT141	Electrodes/Leads and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	16/11/2023	16/11/2023
									A071102	Failure to Interrogate	16/11/2023	16/11/2023
							F1901	Additional Surgery	A070504	Premature Discharge of Battery	16/11/2023	16/11/2023
									A071102	Failure to Interrogate	16/11/2023	16/11/2023
					CT145	Cardiovascular devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	16/11/2023	16/11/2023
									A071102	Failure to Interrogate	16/11/2023	16/11/2023
							F1901	Additional Surgery	A070504	Premature Discharge of Battery	16/11/2023	16/11/2023
									A071102	Failure to Interrogate	16/11/2023	16/11/2023
					CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	16/11/2023	16/11/2023
									A071102	Failure to Interrogate	16/11/2023	16/11/2023
							F1901	Additional Surgery	A070504	Premature Discharge of Battery	16/11/2023	16/11/2023
									A071102	Failure to Interrogate	16/11/2023	16/11/2023
					CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	16/11/2023	16/11/2023
									A071102	Failure to Interrogate	16/11/2023	16/11/2023
							F1901	Additional Surgery	A070504	Premature Discharge of Battery	16/11/2023	16/11/2023
									A071102	Failure to Interrogate	16/11/2023	16/11/2023
					CT233	Surgical	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	16/11/2023	16/11/2023
									A071102	Failure to Interrogate	16/11/2023	16/11/2023
							F1901	Additional Surgery	A070504	Premature Discharge of Battery	16/11/2023	16/11/2023
									A071102	Failure to Interrogate	16/11/2023	16/11/2023
					CT2406	Implantable	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	16/11/2023	16/11/2023
									A071102	Failure to Interrogate	16/11/2023	16/11/2023
							F1901	Additional Surgery	A070504	Premature Discharge of Battery	16/11/2023	16/11/2023
									A071102	Failure to Interrogate	16/11/2023	16/11/2023
					CT2407	Active implantable	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	16/11/2023	16/11/2023
									A071102	Failure to Interrogate	16/11/2023	16/11/2023
							F1901	Additional Surgery	A070504	Premature Discharge of Battery	16/11/2023	16/11/2023
									A071102	Failure to Interrogate	16/11/2023	16/11/2023
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	16/11/2023	16/11/2023
									A071102	Failure to Interrogate	16/11/2023	16/11/2023
							F1901	Additional Surgery	A070504	Premature Discharge of Battery	16/11/2023	16/11/2023
									A071102	Failure to Interrogate	16/11/2023	16/11/2023
					CT2727	Active implant devices/ancillaries	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	16/11/2023	16/11/2023
									A071102	Failure to Interrogate	16/11/2023	16/11/2023
							F1901	Additional Surgery	A070504	Premature Discharge of Battery	16/11/2023	16/11/2023
									A071102	Failure to Interrogate	16/11/2023	16/11/2023
					CT2750	Implantable cardiac electrical stimulators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	16/11/2023	16/11/2023
									A071102	Failure to Interrogate	16/11/2023	16/11/2023
							F1901	Additional Surgery	A070504	Premature Discharge of Battery	16/11/2023	16/11/2023
									A071102	Failure to Interrogate	16/11/2023	16/11/2023
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	16/11/2023	16/11/2023
									A071102	Failure to Interrogate	16/11/2023	16/11/2023
							F1901	Additional Surgery	A070504	Premature Discharge of Battery	16/11/2023	16/11/2023
									A071102	Failure to Interrogate	16/11/2023	16/11/2023
					CT2985	Active	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	16/11/2023	16/11/2023
									A071102	Failure to Interrogate	16/11/2023	16/11/2023
							F1901	Additional Surgery	A070504	Premature Discharge of Battery	16/11/2023	16/11/2023
									A071102	Failure to Interrogate	16/11/2023	16/11/2023
					CT3001	Implantable metal	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	16/11/2023	16/11/2023
									A071102	Failure to Interrogate	16/11/2023	16/11/2023
							F1901	Additional Surgery	A070504	Premature Discharge of Battery	16/11/2023	16/11/2023
									A071102	Failure to Interrogate	16/11/2023	16/11/2023
					CT3002	Implantable synthetic polymer	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	16/11/2023	16/11/2023
									A071102	Failure to Interrogate	16/11/2023	16/11/2023
							F1901	Additional Surgery	A070504	Premature Discharge of Battery	16/11/2023	16/11/2023
									A071102	Failure to Interrogate	16/11/2023	16/11/2023
					CT3006	Implantable inorganic material	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	16/11/2023	16/11/2023

									A071102	Failure to Interrogate	16/11/2023	16/11/2023
							F1901	Additional Surgery	A070504	Premature Discharge of Battery	16/11/2023	16/11/2023
									A071102	Failure to Interrogate	16/11/2023	16/11/2023
					CT314	Electrically powered	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	16/11/2023	16/11/2023
									A071102	Failure to Interrogate	16/11/2023	16/11/2023
							F1901	Additional Surgery	A070504	Premature Discharge of Battery	16/11/2023	16/11/2023
									A071102	Failure to Interrogate	16/11/2023	16/11/2023
					CT321	Long-term surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	16/11/2023	16/11/2023
									A071102	Failure to Interrogate	16/11/2023	16/11/2023
							F1901	Additional Surgery	A070504	Premature Discharge of Battery	16/11/2023	16/11/2023
									A071102	Failure to Interrogate	16/11/2023	16/11/2023
					CT449	Generators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	16/11/2023	16/11/2023
									A071102	Failure to Interrogate	16/11/2023	16/11/2023
							F1901	Additional Surgery	A070504	Premature Discharge of Battery	16/11/2023	16/11/2023
									A071102	Failure to Interrogate	16/11/2023	16/11/2023
					CT981	Single use	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	16/11/2023	16/11/2023
									A071102	Failure to Interrogate	16/11/2023	16/11/2023
							F1901	Additional Surgery	A070504	Premature Discharge of Battery	16/11/2023	16/11/2023
									A071102	Failure to Interrogate	16/11/2023	16/11/2023
					CT983	Surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	16/11/2023	16/11/2023
									A071102	Failure to Interrogate	16/11/2023	16/11/2023
							F1901	Additional Surgery	A070504	Premature Discharge of Battery	16/11/2023	16/11/2023
									A071102	Failure to Interrogate	16/11/2023	16/11/2023
14449	2023		47265	Dual-chamber implantable pacemaker, rate-responsive	CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	19/10/2023	19/10/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	19/10/2023	19/10/2023
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	19/10/2023	19/10/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	19/10/2023	19/10/2023
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	19/10/2023	19/10/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	19/10/2023	19/10/2023
14450	2023	All other reportable incidents	47205	Cardiac pulse generator programmer	CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A11	Computer Software Problem		
									A13	Communication or Transmission Problem		
							F26	No Health Consequences or Impact	A11	Computer Software Problem		
									A13	Communication or Transmission Problem		
14451	2023	All other reportable incidents	47267	Single-chamber implantable pacemaker, rate-responsive	CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
14452	2023	All other reportable incidents	47263	Cardiac resynchronization therapy implantable pacemaker	CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	23/07/2024	23/07/2024
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	23/07/2024	23/07/2024
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	23/07/2024	23/07/2024
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	23/07/2024	23/07/2024
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	23/07/2024	23/07/2024
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	23/07/2024	23/07/2024
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	23/07/2024	23/07/2024
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	23/07/2024	23/07/2024
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	23/07/2024	23/07/2024
14453	2023	All other reportable incidents	47270	Cardiac resynchronization therapy implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
14454	2023	All other reportable incidents	47265	Dual-chamber implantable pacemaker, rate-responsive	CT141	Electrodes/Leads and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070909	Oversensing		
									A0712	Pacing Problem		
									A090801	Signal Artifact/Noise		
							F1001	Absence of Treatment	A070909	Oversensing		
									A0712	Pacing Problem		
									A090801	Signal Artifact/Noise		
					CT145	Cardiovascular devices	E2403	No Clinical Signs, Symptoms or Conditions	A070909	Oversensing		
									A0712	Pacing Problem		
									A090801	Signal Artifact/Noise		
							F1001	Absence of Treatment	A070909	Oversensing		
									A0712	Pacing Problem		
									A090801	Signal Artifact/Noise		
					CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070909	Oversensing		
									A0712	Pacing Problem		
									A090801	Signal Artifact/Noise		
							F1001	Absence of Treatment	A070909	Oversensing		
									A0712	Pacing Problem		
									A090801	Signal Artifact/Noise		
					CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A070909	Oversensing		
									A0712	Pacing Problem		
									A090801	Signal Artifact/Noise		
							F1001	Absence of Treatment	A070909	Oversensing		
									A0712	Pacing Problem		
									A090801	Signal Artifact/Noise		
					CT233	Surgical	E2403	No Clinical Signs, Symptoms or Conditions	A070909	Oversensing		
									A0712	Pacing Problem		
									A090801	Signal Artifact/Noise		
							F1001	Absence of Treatment	A070909	Oversensing		
									A0712	Pacing Problem		
									A090801	Signal Artifact/Noise		
					CT2406	Implantable	E2403	No Clinical Signs, Symptoms or Conditions	A070909	Oversensing		
									A0712	Pacing Problem		
									A090801	Signal Artifact/Noise		
							F1001	Absence of Treatment	A070909	Oversensing		
									A0712	Pacing Problem		
									A090801	Signal Artifact/Noise		
					CT2407	Active implantable	E2403	No Clinical Signs, Symptoms or Conditions	A070909	Oversensing		
									A0712	Pacing Problem		

									A090801	Signal Artifact/Noise		
							F1001	Absence of Treatment	A070909	Oversensing		
									A0712	Pacing Problem		
									A090801	Signal Artifact/Noise		
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A070909	Oversensing		
									A0712	Pacing Problem		
									A090801	Signal Artifact/Noise		
							F1001	Absence of Treatment	A070909	Oversensing		
									A0712	Pacing Problem		
									A090801	Signal Artifact/Noise		
					CT2727	Active implant devices/ancillaries	E2403	No Clinical Signs, Symptoms or Conditions	A070909	Oversensing		
									A0712	Pacing Problem		
									A090801	Signal Artifact/Noise		
							F1001	Absence of Treatment	A070909	Oversensing		
									A0712	Pacing Problem		
									A090801	Signal Artifact/Noise		
					CT2750	Implantable cardiac electrical stimulators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070909	Oversensing		
									A0712	Pacing Problem		
									A090801	Signal Artifact/Noise		
							F1001	Absence of Treatment	A070909	Oversensing		
									A0712	Pacing Problem		
									A090801	Signal Artifact/Noise		
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A070909	Oversensing		
									A0712	Pacing Problem		
									A090801	Signal Artifact/Noise		
							F1001	Absence of Treatment	A070909	Oversensing		
									A0712	Pacing Problem		
									A090801	Signal Artifact/Noise		
					CT2985	Active	E2403	No Clinical Signs, Symptoms or Conditions	A070909	Oversensing		
									A0712	Pacing Problem		
									A090801	Signal Artifact/Noise		
							F1001	Absence of Treatment	A070909	Oversensing		
									A0712	Pacing Problem		
									A090801	Signal Artifact/Noise		
					CT3001	Implantable metal	E2403	No Clinical Signs, Symptoms or Conditions	A070909	Oversensing		
									A0712	Pacing Problem		
									A090801	Signal Artifact/Noise		
							F1001	Absence of Treatment	A070909	Oversensing		
									A0712	Pacing Problem		
									A090801	Signal Artifact/Noise		
					CT3002	Implantable synthetic polymer	E2403	No Clinical Signs, Symptoms or Conditions	A070909	Oversensing		
									A0712	Pacing Problem		
									A090801	Signal Artifact/Noise		
							F1001	Absence of Treatment	A070909	Oversensing		
									A0712	Pacing Problem		
									A090801	Signal Artifact/Noise		
					CT3006	Implantable inorganic material	E2403	No Clinical Signs, Symptoms or Conditions	A070909	Oversensing		
									A0712	Pacing Problem		
									A090801	Signal Artifact/Noise		
							F1001	Absence of Treatment	A070909	Oversensing		
									A0712	Pacing Problem		
									A090801	Signal Artifact/Noise		
					CT314	Electrically powered	E2403	No Clinical Signs, Symptoms or Conditions	A070909	Oversensing		
									A0712	Pacing Problem		
									A090801	Signal Artifact/Noise		
							F1001	Absence of Treatment	A070909	Oversensing		
									A0712	Pacing Problem		
									A090801	Signal Artifact/Noise		
					CT321	Long-term surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A070909	Oversensing		
									A0712	Pacing Problem		
									A090801	Signal Artifact/Noise		
							F1001	Absence of Treatment	A070909	Oversensing		
									A0712	Pacing Problem		
									A090801	Signal Artifact/Noise		
					CT449	Generators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070909	Oversensing		
									A0712	Pacing Problem		
									A090801	Signal Artifact/Noise		
							F1001	Absence of Treatment	A070909	Oversensing		
									A0712	Pacing Problem		
									A090801	Signal Artifact/Noise		
					CT981	Single use	E2403	No Clinical Signs, Symptoms or Conditions	A070909	Oversensing		
									A0712	Pacing Problem		
									A090801	Signal Artifact/Noise		
							F1001	Absence of Treatment	A070909	Oversensing		
									A0712	Pacing Problem		
									A090801	Signal Artifact/Noise		
					CT983	Surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A070909	Oversensing		
									A0712	Pacing Problem		
									A090801	Signal Artifact/Noise		
							F1001	Absence of Treatment	A070909	Oversensing		
									A0712	Pacing Problem		
									A090801	Signal Artifact/Noise		
14455	2023	All other reportable incidents	47270	Cardiac resynchronization therapy implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	12/02/2024	12/02/2024
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	12/02/2024	12/02/2024
							F12	Serious Injury/ Illness/ Impairment	A070504	Premature Discharge of Battery	12/02/2024	12/02/2024
							F1903	Device Explantation	A070504	Premature Discharge of Battery	12/02/2024	12/02/2024
14456	2023	All other reportable incidents	37265	Dual-chamber implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	23/11/2023	23/11/2023
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	23/11/2023	23/11/2023

							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	23/11/2023	23/11/2023
14457	2023		47265	Dual-chamber implantable pacemaker, rate-responsive	CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	27/10/2023	27/10/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	27/10/2023	27/10/2023
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	27/10/2023	27/10/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	27/10/2023	27/10/2023
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	27/10/2023	27/10/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	27/10/2023	27/10/2023
14458	2023	All other reportable incidents	47265	Dual-chamber implantable pacemaker, rate-responsive	CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A0723	Unexpected Mode Switch	14/11/2023	14/11/2023
							F1901	Additional Surgery	A0723	Unexpected Mode Switch	14/11/2023	14/11/2023
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A0723	Unexpected Mode Switch	14/11/2023	14/11/2023
							F1901	Additional Surgery	A0723	Unexpected Mode Switch	14/11/2023	14/11/2023
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A0723	Unexpected Mode Switch	14/11/2023	14/11/2023
							F1901	Additional Surgery	A0723	Unexpected Mode Switch	14/11/2023	14/11/2023
14459	2023		47265	Dual-chamber implantable pacemaker, rate-responsive	CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	15/11/2023	15/11/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	15/11/2023	15/11/2023
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	15/11/2023	15/11/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	15/11/2023	15/11/2023
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	15/11/2023	15/11/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	15/11/2023	15/11/2023
14460	2023		47265	Dual-chamber implantable pacemaker, rate-responsive	CT141	Electrodes/Leads and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	07/11/2023	07/11/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	07/11/2023	07/11/2023
					CT145	Cardiovascular devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	07/11/2023	07/11/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	07/11/2023	07/11/2023
					CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	07/11/2023	07/11/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	07/11/2023	07/11/2023
					CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	07/11/2023	07/11/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	07/11/2023	07/11/2023
					CT233	Surgical	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	07/11/2023	07/11/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	07/11/2023	07/11/2023
					CT2406	Implantable	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	07/11/2023	07/11/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	07/11/2023	07/11/2023
					CT2407	Active implantable	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	07/11/2023	07/11/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	07/11/2023	07/11/2023
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	07/11/2023	07/11/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	07/11/2023	07/11/2023
					CT2727	Active implant devices/ancillaries	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	07/11/2023	07/11/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	07/11/2023	07/11/2023
					CT2750	Implantable cardiac electrical stimulators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	07/11/2023	07/11/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	07/11/2023	07/11/2023
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	07/11/2023	07/11/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	07/11/2023	07/11/2023
					CT2985	Active	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	07/11/2023	07/11/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	07/11/2023	07/11/2023
					CT3001	Implantable metal	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	07/11/2023	07/11/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	07/11/2023	07/11/2023
					CT3002	Implantable synthetic polymer	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	07/11/2023	07/11/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	07/11/2023	07/11/2023
					CT3006	Implantable inorganic material	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	07/11/2023	07/11/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	07/11/2023	07/11/2023
					CT314	Electrically powered	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	07/11/2023	07/11/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	07/11/2023	07/11/2023
					CT321	Long-term surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	07/11/2023	07/11/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	07/11/2023	07/11/2023
					CT449	Generators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	07/11/2023	07/11/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	07/11/2023	07/11/2023
					CT981	Single use	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	07/11/2023	07/11/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	07/11/2023	07/11/2023
					CT983	Surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	07/11/2023	07/11/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	07/11/2023	07/11/2023
14461	2023		47265	Dual-chamber implantable pacemaker, rate-responsive	CT141	Electrodes/Leads and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	07/11/2023	07/11/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	07/11/2023	07/11/2023
					CT145	Cardiovascular devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	07/11/2023	07/11/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	07/11/2023	07/11/2023
					CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	07/11/2023	07/11/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	07/11/2023	07/11/2023
					CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	07/11/2023	07/11/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	07/11/2023	07/11/2023
					CT233	Surgical	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	07/11/2023	07/11/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	07/11/2023	07/11/2023
					CT2406	Implantable	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	07/11/2023	07/11/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	07/11/2023	07/11/2023
					CT2407	Active implantable	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	07/11/2023	07/11/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	07/11/2023	07/11/2023
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	07/11/2023	07/11/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	07/11/2023	07/11/2023
					CT2727	Active implant devices/ancillaries	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	07/11/2023	07/11/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	07/11/2023	07/11/2023
					CT2750	Implantable cardiac electrical stimulators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	07/11/2023	07/11/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	07/11/2023	07/11/2023
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	07/11/2023	07/11/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	07/11/2023	07/11/2023
					CT2985	Active	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	07/11/2023	07/11/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	07/11/2023	07/11/2023
					CT3001	Implantable metal	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	07/11/2023	07/11/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	07/11/2023	07/11/2023
					CT3002	Implantable synthetic polymer	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	07/11/2023	07/11/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	07/11/2023	07/11/2023
					CT3006	Implantable inorganic material	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	07/11/2023	07/11/2023

							F1901	Additional Surgery	A24	No Apparent Device Problem	07/11/2023	07/11/2023
					CT314	Electrically powered	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	07/11/2023	07/11/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	07/11/2023	07/11/2023
					CT321	Long-term surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	07/11/2023	07/11/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	07/11/2023	07/11/2023
					CT449	Generators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	07/11/2023	07/11/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	07/11/2023	07/11/2023
					CT981	Single use	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	07/11/2023	07/11/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	07/11/2023	07/11/2023
					CT983	Surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	07/11/2023	07/11/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	07/11/2023	07/11/2023
14462	2023	All other reportable incidents	63188	Left-ventricular wireless pacing system sensor/transmitter	CT145	Cardiovascular devices			A0708	Power Problem		
					CT1577	Wireless monitoring system transmitters			A0708	Power Problem		
					CT1684	Cardiac pacemakers and associated devices			A0708	Power Problem		
					CT2727	Active implant devices/ancillaries			A0708	Power Problem		
					CT2750	Implantable cardiac electrical stimulators and associated devices			A0708	Power Problem		
					CT2885	Physiological monitoring/measuring system transmitters/receivers/transceivers			A0708	Power Problem		
14463	2023		47265	Dual-chamber implantable pacemaker, rate-responsive	CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	12/09/2023	12/09/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	12/09/2023	12/09/2023
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	12/09/2023	12/09/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	12/09/2023	12/09/2023
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	12/09/2023	12/09/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	12/09/2023	12/09/2023
14464	2023	All other reportable incidents	35852	Single-chamber implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	11/02/2024	11/02/2024
									A0908	Incorrect, Inadequate or Imprecise Result or Readings	11/02/2024	11/02/2024
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	11/02/2024	11/02/2024
									A0908	Incorrect, Inadequate or Imprecise Result or Readings	11/02/2024	11/02/2024
							F1903	Device Explantation	A070504	Premature Discharge of Battery	11/02/2024	11/02/2024
									A0908	Incorrect, Inadequate or Imprecise Result or Readings	11/02/2024	11/02/2024
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	11/02/2024	11/02/2024
									A0908	Incorrect, Inadequate or Imprecise Result or Readings	11/02/2024	11/02/2024
14465	2023	All other reportable incidents	35852	Single-chamber implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A0709	Device Sensing Problem		
									A070910	Undersensing		
									A1301	Failure to Read Input Signal		
							F2203	Imaging Required	A0709	Device Sensing Problem		
									A070910	Undersensing		
									A1301	Failure to Read Input Signal		
							F26	No Health Consequences or Impact	A0709	Device Sensing Problem		
									A070910	Undersensing		
									A1301	Failure to Read Input Signal		
14466	2023	All other reportable incidents	47265	Dual-chamber implantable pacemaker, rate-responsive	CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
14467	2023	All other reportable incidents	47267	Single-chamber implantable pacemaker, rate-responsive	CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	13/11/2023	13/11/2023
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	13/11/2023	13/11/2023
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	13/11/2023	13/11/2023
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	13/11/2023	13/11/2023
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	13/11/2023	13/11/2023
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	13/11/2023	13/11/2023
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	13/11/2023	13/11/2023
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	13/11/2023	13/11/2023
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	13/11/2023	13/11/2023
14468	2023		47263	Cardiac resynchronization therapy implantable pacemaker	CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	12/08/2020	12/08/2020
							F1901	Additional Surgery	A070504	Premature Discharge of Battery	12/08/2020	12/08/2020
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	12/08/2020	12/08/2020
							F1901	Additional Surgery	A070504	Premature Discharge of Battery	12/08/2020	12/08/2020
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	12/08/2020	12/08/2020
							F1901	Additional Surgery	A070504	Premature Discharge of Battery	12/08/2020	12/08/2020
14469	2023		47265	Dual-chamber implantable pacemaker, rate-responsive	CT141	Electrodes/Leads and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	17/10/2023	17/10/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	17/10/2023	17/10/2023
					CT145	Cardiovascular devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	17/10/2023	17/10/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	17/10/2023	17/10/2023
					CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	17/10/2023	17/10/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	17/10/2023	17/10/2023
					CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	17/10/2023	17/10/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	17/10/2023	17/10/2023
					CT233	Surgical	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	17/10/2023	17/10/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	17/10/2023	17/10/2023
					CT2406	Implantable	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	17/10/2023	17/10/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	17/10/2023	17/10/2023
					CT2407	Active implantable	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	17/10/2023	17/10/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	17/10/2023	17/10/2023
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	17/10/2023	17/10/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	17/10/2023	17/10/2023

					CT2727	Active implant devices/ancillaries	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	17/10/2023	17/10/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	17/10/2023	17/10/2023
					CT2750	Implantable cardiac electrical stimulators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	17/10/2023	17/10/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	17/10/2023	17/10/2023
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	17/10/2023	17/10/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	17/10/2023	17/10/2023
					CT2985	Active	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	17/10/2023	17/10/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	17/10/2023	17/10/2023
					CT3001	Implantable metal	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	17/10/2023	17/10/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	17/10/2023	17/10/2023
					CT3002	Implantable synthetic polymer	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	17/10/2023	17/10/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	17/10/2023	17/10/2023
					CT3006	Implantable inorganic material	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	17/10/2023	17/10/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	17/10/2023	17/10/2023
					CT314	Electrically powered	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	17/10/2023	17/10/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	17/10/2023	17/10/2023
					CT321	Long-term surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	17/10/2023	17/10/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	17/10/2023	17/10/2023
					CT449	Generators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	17/10/2023	17/10/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	17/10/2023	17/10/2023
					CT981	Single use	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	17/10/2023	17/10/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	17/10/2023	17/10/2023
					CT983	Surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	17/10/2023	17/10/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	17/10/2023	17/10/2023
14470	2023		47265	Dual-chamber implantable pacemaker, rate-responsive	CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A25	No Apparent Adverse Event	19/09/2023	19/09/2023
							F26	No Health Consequences or Impact	A25	No Apparent Adverse Event	19/09/2023	19/09/2023
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A25	No Apparent Adverse Event	19/09/2023	19/09/2023
							F26	No Health Consequences or Impact	A25	No Apparent Adverse Event	19/09/2023	19/09/2023
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A25	No Apparent Adverse Event	19/09/2023	19/09/2023
							F26	No Health Consequences or Impact	A25	No Apparent Adverse Event	19/09/2023	19/09/2023
14471	2023	All other reportable incidents	35852	Single-chamber implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F1901	Additional Surgery	A070504	Premature Discharge of Battery		
14472	2023		47265	Dual-chamber implantable pacemaker, rate-responsive	CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	17/11/2023	17/11/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	17/11/2023	17/11/2023
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	17/11/2023	17/11/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	17/11/2023	17/11/2023
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	17/11/2023	17/11/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	17/11/2023	17/11/2023
14473	2023	All other reportable incidents	17882	Physiologic-monitoring defibrillation system	CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A0722	Impedance Problem		
							F27	Problem Identified During Non-Clinical Procedure	A0722	Impedance Problem		
14474	2023	All other reportable incidents	36241	Pacemaker/defibrillator lead	CT1684	Cardiac pacemakers and associated devices	E011903	Syncope/Fainting	A072201	High Impedance		
							F08	Hospitalization or Prolonged Hospitalization	A072201	High Impedance		
14475	2023	All other reportable incidents	36241	Pacemaker/defibrillator lead	CT1684	Cardiac pacemakers and associated devices	E011903	Syncope/Fainting	A072201	High Impedance		
							F08	Hospitalization or Prolonged Hospitalization	A072201	High Impedance		
14476	2023	All other reportable incidents	47265	Dual-chamber implantable pacemaker, rate-responsive	CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	13/11/2023	13/11/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	13/11/2023	13/11/2023
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	13/11/2023	13/11/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	13/11/2023	13/11/2023
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	13/11/2023	13/11/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	13/11/2023	13/11/2023
14477	2023	All other reportable incidents	47265	Dual-chamber implantable pacemaker, rate-responsive	CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem	15/02/2024	15/02/2024
									A070504	Premature Discharge of Battery	15/02/2024	15/02/2024
							F08	Hospitalization or Prolonged Hospitalization	A0705	Battery Problem	15/02/2024	15/02/2024
									A070504	Premature Discharge of Battery	15/02/2024	15/02/2024
							F1905	Device Revision or Replacement	A0705	Battery Problem	15/02/2024	15/02/2024
									A070504	Premature Discharge of Battery	15/02/2024	15/02/2024
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem	15/02/2024	15/02/2024
									A070504	Premature Discharge of Battery	15/02/2024	15/02/2024
							F08	Hospitalization or Prolonged Hospitalization	A0705	Battery Problem	15/02/2024	15/02/2024
									A070504	Premature Discharge of Battery	15/02/2024	15/02/2024
							F1905	Device Revision or Replacement	A0705	Battery Problem	15/02/2024	15/02/2024
									A070504	Premature Discharge of Battery	15/02/2024	15/02/2024
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem	15/02/2024	15/02/2024
									A070504	Premature Discharge of Battery	15/02/2024	15/02/2024
							F08	Hospitalization or Prolonged Hospitalization	A0705	Battery Problem	15/02/2024	15/02/2024
									A070504	Premature Discharge of Battery	15/02/2024	15/02/2024
							F1905	Device Revision or Replacement	A0705	Battery Problem	15/02/2024	15/02/2024
									A070504	Premature Discharge of Battery	15/02/2024	15/02/2024
14478	2023		35852	Single-chamber implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E2401	Insufficient Information	A070910	Undersensing		
							F02	Death	A070910	Undersensing		
14479	2023	All other reportable incidents	47265	Dual-chamber implantable pacemaker, rate-responsive	CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A070101	Failure to Capture		
							F23	Unexpected Medical Intervention	A070101	Failure to Capture		
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A070101	Failure to Capture		
							F23	Unexpected Medical Intervention	A070101	Failure to Capture		
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A070101	Failure to Capture		
							F23	Unexpected Medical Intervention	A070101	Failure to Capture		
14480	2023	All other reportable incidents	17882	Physiologic-monitoring defibrillation system	CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A0719	Unexpected Shutdown		
							F26	No Health Consequences or Impact	A0719	Unexpected Shutdown		
14481	2023	All other reportable incidents	17882	Physiologic-monitoring defibrillation system	CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070903	Failure to Analyze Signal		
							F26	No Health Consequences or Impact	A070903	Failure to Analyze Signal		
14482	2023	All other reportable incidents	35852	Single-chamber implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	09/01/2024	09/01/2024
									A0908	Incorrect, Inadequate or Imprecise Result or Readings	09/01/2024	09/01/2024
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	09/01/2024	09/01/2024
									A0908	Incorrect, Inadequate or Imprecise Result or Readings	09/01/2024	09/01/2024
							F1903	Device Explantation	A070504	Premature Discharge of Battery	09/01/2024	09/01/2024

									A0908	Incorrect, Inadequate or Imprecise Result or Readings	09/01/2024	09/01/2024	
14483	2023	All other reportable incidents	63188	Left-ventricular wireless pacing system sensor/transmitter	CT145	Cardiovascular devices			A0708	Power Problem			
					CT1577	Wireless monitoring system transmitters			A0708	Power Problem			
					CT1684	Cardiac pacemakers and associated devices			A0708	Power Problem			
					CT2727	Active implant devices/ancillaries			A0708	Power Problem			
					CT2750	Implantable cardiac electrical stimulators and associated devices			A0708	Power Problem			
					CT2885	Physiological monitoring/measuring system transmitters/receivers/transceivers			A0708	Power Problem			
14484	2023	All other reportable incidents	47265	Dual-chamber implantable pacemaker, rate-responsive	CT141	Electrodes/Leads and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem			
									A070504	Premature Discharge of Battery			
							F26	No Health Consequences or Impact	A0705	Battery Problem			
									A070504	Premature Discharge of Battery			
					CT145	Cardiovascular devices	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem			
									A070504	Premature Discharge of Battery			
							F26	No Health Consequences or Impact	A0705	Battery Problem			
									A070504	Premature Discharge of Battery			
					CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem			
									A070504	Premature Discharge of Battery			
							F26	No Health Consequences or Impact	A0705	Battery Problem			
									A070504	Premature Discharge of Battery			
					CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem			
									A070504	Premature Discharge of Battery			
							F26	No Health Consequences or Impact	A0705	Battery Problem			
									A070504	Premature Discharge of Battery			
					CT233	Surgical	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem			
									A070504	Premature Discharge of Battery			
							F26	No Health Consequences or Impact	A0705	Battery Problem			
									A070504	Premature Discharge of Battery			
					CT2406	Implantable	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem			
									A070504	Premature Discharge of Battery			
							F26	No Health Consequences or Impact	A0705	Battery Problem			
									A070504	Premature Discharge of Battery			
					CT2407	Active implantable	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem			
									A070504	Premature Discharge of Battery			
							F26	No Health Consequences or Impact	A0705	Battery Problem			
									A070504	Premature Discharge of Battery			
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem			
									A070504	Premature Discharge of Battery			
							F26	No Health Consequences or Impact	A0705	Battery Problem			
									A070504	Premature Discharge of Battery			
					CT2727	Active implant devices/ancillaries	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem			
									A070504	Premature Discharge of Battery			
							F26	No Health Consequences or Impact	A0705	Battery Problem			
									A070504	Premature Discharge of Battery			
					CT2750	Implantable cardiac electrical stimulators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem			
									A070504	Premature Discharge of Battery			
							F26	No Health Consequences or Impact	A0705	Battery Problem			
									A070504	Premature Discharge of Battery			
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem			
									A070504	Premature Discharge of Battery			
							F26	No Health Consequences or Impact	A0705	Battery Problem			
									A070504	Premature Discharge of Battery			
					CT2985	Active	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem			
									A070504	Premature Discharge of Battery			
							F26	No Health Consequences or Impact	A0705	Battery Problem			
									A070504	Premature Discharge of Battery			
					CT3001	Implantable metal	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem			
									A070504	Premature Discharge of Battery			
							F26	No Health Consequences or Impact	A0705	Battery Problem			
									A070504	Premature Discharge of Battery			
					CT3002	Implantable synthetic polymer	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem			
									A070504	Premature Discharge of Battery			
							F26	No Health Consequences or Impact	A0705	Battery Problem			
									A070504	Premature Discharge of Battery			
					CT3006	Implantable inorganic material	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem			
									A070504	Premature Discharge of Battery			
							F26	No Health Consequences or Impact	A0705	Battery Problem			
									A070504	Premature Discharge of Battery			
					CT314	Electrically powered	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem			
									A070504	Premature Discharge of Battery			
							F26	No Health Consequences or Impact	A0705	Battery Problem			
									A070504	Premature Discharge of Battery			
					CT321	Long-term surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem			
									A070504	Premature Discharge of Battery			
							F26	No Health Consequences or Impact	A0705	Battery Problem			
									A070504	Premature Discharge of Battery			
					CT449	Generators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem			
									A070504	Premature Discharge of Battery			
							F26	No Health Consequences or Impact	A0705	Battery Problem			
									A070504	Premature Discharge of Battery			
					CT981	Single use	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem			
									A070504	Premature Discharge of Battery			
							F26	No Health Consequences or Impact	A0705	Battery Problem			
									A070504	Premature Discharge of Battery			
					CT983	Surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem			
									A070504	Premature Discharge of Battery			

							F26	No Health Consequences or Impact	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
14485	2023	All other reportable incidents	47265	Dual-chamber implantable pacemaker, rate-responsive	CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	14/11/2023	14/11/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	14/11/2023	14/11/2023
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	14/11/2023	14/11/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	14/11/2023	14/11/2023
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	14/11/2023	14/11/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	14/11/2023	14/11/2023
14486	2023	All other reportable incidents	47265	Dual-chamber implantable pacemaker, rate-responsive	CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	21/11/2023	21/11/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	21/11/2023	21/11/2023
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	21/11/2023	21/11/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	21/11/2023	21/11/2023
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	21/11/2023	21/11/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	21/11/2023	21/11/2023
14487	2023		47265	Dual-chamber implantable pacemaker, rate-responsive	CT141	Electrodes/Leads and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	21/11/2023	21/11/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	21/11/2023	21/11/2023
					CT145	Cardiovascular devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	21/11/2023	21/11/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	21/11/2023	21/11/2023
					CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	21/11/2023	21/11/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	21/11/2023	21/11/2023
					CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	21/11/2023	21/11/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	21/11/2023	21/11/2023
					CT233	Surgical	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	21/11/2023	21/11/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	21/11/2023	21/11/2023
					CT2406	Implantable	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	21/11/2023	21/11/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	21/11/2023	21/11/2023
					CT2407	Active implantable	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	21/11/2023	21/11/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	21/11/2023	21/11/2023
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	21/11/2023	21/11/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	21/11/2023	21/11/2023
					CT2727	Active implant devices/ancillaries	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	21/11/2023	21/11/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	21/11/2023	21/11/2023
					CT2750	Implantable cardiac electrical stimulators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	21/11/2023	21/11/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	21/11/2023	21/11/2023
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	21/11/2023	21/11/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	21/11/2023	21/11/2023
					CT2985	Active	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	21/11/2023	21/11/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	21/11/2023	21/11/2023
					CT3001	Implantable metal	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	21/11/2023	21/11/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	21/11/2023	21/11/2023
					CT3002	Implantable synthetic polymer	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	21/11/2023	21/11/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	21/11/2023	21/11/2023
					CT3006	Implantable inorganic material	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	21/11/2023	21/11/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	21/11/2023	21/11/2023
					CT314	Electrically powered	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	21/11/2023	21/11/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	21/11/2023	21/11/2023
					CT321	Long-term surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	21/11/2023	21/11/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	21/11/2023	21/11/2023
					CT449	Generators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	21/11/2023	21/11/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	21/11/2023	21/11/2023
					CT981	Single use	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	21/11/2023	21/11/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	21/11/2023	21/11/2023
					CT983	Surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	21/11/2023	21/11/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	21/11/2023	21/11/2023
14488	2023		47265	Dual-chamber implantable pacemaker, rate-responsive	CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	21/11/2023	21/11/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	21/11/2023	21/11/2023
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	21/11/2023	21/11/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	21/11/2023	21/11/2023
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	21/11/2023	21/11/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	21/11/2023	21/11/2023
14489	2023	All other reportable incidents	37265	Dual-chamber implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A072201	High Impedance		
									A0908	Incorrect, Inadequate or Imprecise Result or Readings		
							F26	No Health Consequences or Impact	A072201	High Impedance		
									A0908	Incorrect, Inadequate or Imprecise Result or Readings		
14490	2023	All other reportable incidents	35852	Single-chamber implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E2104	Electric Shock	A0712	Pacing Problem		
									A071302	Inappropriate/Inadequate Shock/Stimulation		
							F23	Unexpected Medical Intervention	A0712	Pacing Problem		
									A071302	Inappropriate/Inadequate Shock/Stimulation		
14491	2023	All other reportable incidents	60887	Intracardiac pacemaker programming interface unit	CT1684	Cardiac pacemakers and associated devices	E0112	Dizziness	A0908	Incorrect, Inadequate or Imprecise Result or Readings		
							F26	No Health Consequences or Impact	A0908	Incorrect, Inadequate or Imprecise Result or Readings		
14492	2023		47265	Dual-chamber implantable pacemaker, rate-responsive	CT141	Electrodes/Leads and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	09/11/2023	09/11/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	09/11/2023	09/11/2023
					CT145	Cardiovascular devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	09/11/2023	09/11/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	09/11/2023	09/11/2023
					CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	09/11/2023	09/11/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	09/11/2023	09/11/2023
					CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	09/11/2023	09/11/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	09/11/2023	09/11/2023
					CT233	Surgical	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	09/11/2023	09/11/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	09/11/2023	09/11/2023
					CT2406	Implantable	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	09/11/2023	09/11/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	09/11/2023	09/11/2023
					CT2407	Active implantable	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	09/11/2023	09/11/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	09/11/2023	09/11/2023
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	09/11/2023	09/11/2023

						F1901	Additional Surgery	A24	No Apparent Device Problem	09/11/2023	09/11/2023
				CT2727	Active implant devices/ancillaries	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	09/11/2023	09/11/2023
						F1901	Additional Surgery	A24	No Apparent Device Problem	09/11/2023	09/11/2023
				CT2750	Implantable cardiac electrical stimulators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	09/11/2023	09/11/2023
						F1901	Additional Surgery	A24	No Apparent Device Problem	09/11/2023	09/11/2023
				CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	09/11/2023	09/11/2023
						F1901	Additional Surgery	A24	No Apparent Device Problem	09/11/2023	09/11/2023
				CT2985	Active	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	09/11/2023	09/11/2023
						F1901	Additional Surgery	A24	No Apparent Device Problem	09/11/2023	09/11/2023
				CT3001	Implantable metal	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	09/11/2023	09/11/2023
						F1901	Additional Surgery	A24	No Apparent Device Problem	09/11/2023	09/11/2023
				CT3002	Implantable synthetic polymer	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	09/11/2023	09/11/2023
						F1901	Additional Surgery	A24	No Apparent Device Problem	09/11/2023	09/11/2023
				CT3006	Implantable inorganic material	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	09/11/2023	09/11/2023
						F1901	Additional Surgery	A24	No Apparent Device Problem	09/11/2023	09/11/2023
				CT314	Electrically powered	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	09/11/2023	09/11/2023
						F1901	Additional Surgery	A24	No Apparent Device Problem	09/11/2023	09/11/2023
				CT321	Long-term surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	09/11/2023	09/11/2023
						F1901	Additional Surgery	A24	No Apparent Device Problem	09/11/2023	09/11/2023
				CT449	Generators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	09/11/2023	09/11/2023
						F1901	Additional Surgery	A24	No Apparent Device Problem	09/11/2023	09/11/2023
				CT981	Single use	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	09/11/2023	09/11/2023
						F1901	Additional Surgery	A24	No Apparent Device Problem	09/11/2023	09/11/2023
				CT983	Surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	09/11/2023	09/11/2023
						F1901	Additional Surgery	A24	No Apparent Device Problem	09/11/2023	09/11/2023
14493	2023	All other reportable incidents	47265	Dual-chamber implantable pacemaker, rate-responsive	CT141	Electrodes/Leads and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A0712	Pacing Problem		
							F08	Hospitalization or Prolonged Hospitalization	A0712	Pacing Problem		
					CT145	Cardiovascular devices	E2403	No Clinical Signs, Symptoms or Conditions	A0712	Pacing Problem		
							F08	Hospitalization or Prolonged Hospitalization	A0712	Pacing Problem		
					CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A0712	Pacing Problem		
							F08	Hospitalization or Prolonged Hospitalization	A0712	Pacing Problem		
					CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A0712	Pacing Problem		
							F08	Hospitalization or Prolonged Hospitalization	A0712	Pacing Problem		
					CT233	Surgical	E2403	No Clinical Signs, Symptoms or Conditions	A0712	Pacing Problem		
							F08	Hospitalization or Prolonged Hospitalization	A0712	Pacing Problem		
					CT2406	Implantable	E2403	No Clinical Signs, Symptoms or Conditions	A0712	Pacing Problem		
							F08	Hospitalization or Prolonged Hospitalization	A0712	Pacing Problem		
					CT2407	Active implantable	E2403	No Clinical Signs, Symptoms or Conditions	A0712	Pacing Problem		
							F08	Hospitalization or Prolonged Hospitalization	A0712	Pacing Problem		
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A0712	Pacing Problem		
							F08	Hospitalization or Prolonged Hospitalization	A0712	Pacing Problem		
					CT2727	Active implant devices/ancillaries	E2403	No Clinical Signs, Symptoms or Conditions	A0712	Pacing Problem		
							F08	Hospitalization or Prolonged Hospitalization	A0712	Pacing Problem		
					CT2750	Implantable cardiac electrical stimulators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A0712	Pacing Problem		
							F08	Hospitalization or Prolonged Hospitalization	A0712	Pacing Problem		
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A0712	Pacing Problem		
							F08	Hospitalization or Prolonged Hospitalization	A0712	Pacing Problem		
					CT2985	Active	E2403	No Clinical Signs, Symptoms or Conditions	A0712	Pacing Problem		
							F08	Hospitalization or Prolonged Hospitalization	A0712	Pacing Problem		
					CT3001	Implantable metal	E2403	No Clinical Signs, Symptoms or Conditions	A0712	Pacing Problem		
							F08	Hospitalization or Prolonged Hospitalization	A0712	Pacing Problem		
					CT3002	Implantable synthetic polymer	E2403	No Clinical Signs, Symptoms or Conditions	A0712	Pacing Problem		
							F08	Hospitalization or Prolonged Hospitalization	A0712	Pacing Problem		
					CT3006	Implantable inorganic material	E2403	No Clinical Signs, Symptoms or Conditions	A0712	Pacing Problem		
							F08	Hospitalization or Prolonged Hospitalization	A0712	Pacing Problem		
					CT314	Electrically powered	E2403	No Clinical Signs, Symptoms or Conditions	A0712	Pacing Problem		
							F08	Hospitalization or Prolonged Hospitalization	A0712	Pacing Problem		
					CT321	Long-term surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A0712	Pacing Problem		
							F08	Hospitalization or Prolonged Hospitalization	A0712	Pacing Problem		
					CT449	Generators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A0712	Pacing Problem		
							F08	Hospitalization or Prolonged Hospitalization	A0712	Pacing Problem		
					CT981	Single use	E2403	No Clinical Signs, Symptoms or Conditions	A0712	Pacing Problem		
							F08	Hospitalization or Prolonged Hospitalization	A0712	Pacing Problem		
					CT983	Surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A0712	Pacing Problem		
							F08	Hospitalization or Prolonged Hospitalization	A0712	Pacing Problem		
14494	2023	All other reportable incidents	17882	Physiologic-monitoring defibrillation system	CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A0719	Unexpected Shutdown		
							F27	Problem Identified During Non-Clinical Procedure	A0719	Unexpected Shutdown		
14495	2023	All other reportable incidents	35852	Single-chamber implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E2104	Electric Shock	A0709	Device Sensing Problem		
									A070909	Oversensing		
									A071302	Inappropriate/Inadequate Shock/Stimulation		
									A1301	Failure to Read Input Signal		
							F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A0709	Device Sensing Problem		
									A070909	Oversensing		
									A071302	Inappropriate/Inadequate Shock/Stimulation		
									A1301	Failure to Read Input Signal		
14496	2023	All other reportable incidents	35852	Single-chamber implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E2104	Electric Shock	A070909	Oversensing		
									A071302	Inappropriate/Inadequate Shock/Stimulation		
							F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A070909	Oversensing		
									A071302	Inappropriate/Inadequate Shock/Stimulation		
							F13	Misdiagnosis/ Misclassification	A070909	Oversensing		
									A071302	Inappropriate/Inadequate Shock/Stimulation		
14497	2023	All other reportable incidents	47265	Dual-chamber implantable pacemaker, rate-responsive	CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A0908	Incorrect, Inadequate or Imprecise Result or Readings	21/11/2023	21/11/2023
							F08	Hospitalization or Prolonged Hospitalization	A0908	Incorrect, Inadequate or Imprecise Result or Readings	21/11/2023	21/11/2023
							F1903	Device Explantation	A0908	Incorrect, Inadequate or Imprecise Result or Readings	21/11/2023	21/11/2023
							F1905	Device Revision or Replacement	A0908	Incorrect, Inadequate or Imprecise Result or Readings	21/11/2023	21/11/2023

					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A0908	Incorrect, Inadequate or Imprecise Result or Readings	21/11/2023	21/11/2023
							F08	Hospitalization or Prolonged Hospitalization	A0908	Incorrect, Inadequate or Imprecise Result or Readings	21/11/2023	21/11/2023
							F1903	Device Explantation	A0908	Incorrect, Inadequate or Imprecise Result or Readings	21/11/2023	21/11/2023
							F1905	Device Revision or Replacement	A0908	Incorrect, Inadequate or Imprecise Result or Readings	21/11/2023	21/11/2023
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A0908	Incorrect, Inadequate or Imprecise Result or Readings	21/11/2023	21/11/2023
							F08	Hospitalization or Prolonged Hospitalization	A0908	Incorrect, Inadequate or Imprecise Result or Readings	21/11/2023	21/11/2023
							F1903	Device Explantation	A0908	Incorrect, Inadequate or Imprecise Result or Readings	21/11/2023	21/11/2023
							F1905	Device Revision or Replacement	A0908	Incorrect, Inadequate or Imprecise Result or Readings	21/11/2023	21/11/2023
14498	2023	All other reportable incidents	47263	Cardiac resynchronization therapy implantable pacemaker	CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A070101	Failure to Capture		
							F26	No Health Consequences or Impact	A070101	Failure to Capture		
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A070101	Failure to Capture		
							F26	No Health Consequences or Impact	A070101	Failure to Capture		
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A070101	Failure to Capture		
							F26	No Health Consequences or Impact	A070101	Failure to Capture		
14499	2023		47265	Dual-chamber implantable pacemaker, rate-responsive	CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	12/11/2023	12/11/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	12/11/2023	12/11/2023
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	12/11/2023	12/11/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	12/11/2023	12/11/2023
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	12/11/2023	12/11/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	12/11/2023	12/11/2023
14500	2023	All other reportable incidents	37265	Dual-chamber implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E210401	Shock from Patient Lead(s)	A0712	Pacing Problem	17/11/2023	17/11/2023
									A071302	Inappropriate/Inadequate Shock/Stimulation	17/11/2023	17/11/2023
									A16	Protective Measures Problem	17/11/2023	17/11/2023
							F08	Hospitalization or Prolonged Hospitalization	A0712	Pacing Problem	17/11/2023	17/11/2023
									A071302	Inappropriate/Inadequate Shock/Stimulation	17/11/2023	17/11/2023
									A16	Protective Measures Problem	17/11/2023	17/11/2023
							F1203	Life Threatening Illness or Injury	A0712	Pacing Problem	17/11/2023	17/11/2023
									A071302	Inappropriate/Inadequate Shock/Stimulation	17/11/2023	17/11/2023
									A16	Protective Measures Problem	17/11/2023	17/11/2023
							F1903	Device Explantation	A0712	Pacing Problem	17/11/2023	17/11/2023
									A071302	Inappropriate/Inadequate Shock/Stimulation	17/11/2023	17/11/2023
									A16	Protective Measures Problem	17/11/2023	17/11/2023
14501	2023	All other reportable incidents	35852	Single-chamber implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	12/10/2023	12/10/2023
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	12/10/2023	12/10/2023
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	12/10/2023	12/10/2023
14502	2023	All other reportable incidents	37265	Dual-chamber implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A070602	Delayed Charge Time	26/09/2024	26/09/2024
									A0908	Incorrect, Inadequate or Imprecise Result or Readings	26/09/2024	26/09/2024
							F08	Hospitalization or Prolonged Hospitalization	A070602	Delayed Charge Time	26/09/2024	26/09/2024
									A0908	Incorrect, Inadequate or Imprecise Result or Readings	26/09/2024	26/09/2024
							F1001	Absence of Treatment	A070602	Delayed Charge Time	26/09/2024	26/09/2024
									A0908	Incorrect, Inadequate or Imprecise Result or Readings	26/09/2024	26/09/2024
							F1905	Device Revision or Replacement	A070602	Delayed Charge Time	26/09/2024	26/09/2024
									A0908	Incorrect, Inadequate or Imprecise Result or Readings	26/09/2024	26/09/2024
14503	2023	All other reportable incidents	47263	Cardiac resynchronization therapy implantable pacemaker	CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A070102	High Capture Threshold		
									A070909	Oversensing		
									A0712	Pacing Problem		
									A072201	High Impedance		
									A090801	Signal Artifact/Noise		
							F1001	Absence of Treatment	A070102	High Capture Threshold		
									A070909	Oversensing		
									A0712	Pacing Problem		
									A072201	High Impedance		
									A090801	Signal Artifact/Noise		
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A070102	High Capture Threshold		
									A070909	Oversensing		
									A0712	Pacing Problem		
									A072201	High Impedance		
									A090801	Signal Artifact/Noise		
							F1001	Absence of Treatment	A070102	High Capture Threshold		
									A070909	Oversensing		
									A0712	Pacing Problem		
									A072201	High Impedance		
									A090801	Signal Artifact/Noise		
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A070102	High Capture Threshold		
									A070909	Oversensing		
									A0712	Pacing Problem		
									A072201	High Impedance		
									A090801	Signal Artifact/Noise		
							F1001	Absence of Treatment	A070102	High Capture Threshold		
									A070909	Oversensing		
									A0712	Pacing Problem		
									A072201	High Impedance		
									A090801	Signal Artifact/Noise		
14504	2023		17882	Physiologic-monitoring defibrillation system	CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
14505	2023	All other reportable incidents	47265	Dual-chamber implantable pacemaker, rate-responsive	CT141	Electrodes/Leads and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	29/08/2023	29/08/2023
									A071102	Failure to Interrogate	29/08/2023	29/08/2023
									A071204	No Pacing	29/08/2023	29/08/2023
							F1901	Additional Surgery	A070504	Premature Discharge of Battery	29/08/2023	29/08/2023

								A071102	Failure to Interrogate	29/08/2023	29/08/2023
								A071204	No Pacing	29/08/2023	29/08/2023
					CT145	Cardiovascular devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	29/08/2023	29/08/2023
									A071102	Failure to Interrogate	29/08/2023	29/08/2023
									A071204	No Pacing	29/08/2023	29/08/2023
							F1901	Additional Surgery	A070504	Premature Discharge of Battery	29/08/2023	29/08/2023
									A071102	Failure to Interrogate	29/08/2023	29/08/2023
									A071204	No Pacing	29/08/2023	29/08/2023
					CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	29/08/2023	29/08/2023
									A071102	Failure to Interrogate	29/08/2023	29/08/2023
									A071204	No Pacing	29/08/2023	29/08/2023
							F1901	Additional Surgery	A070504	Premature Discharge of Battery	29/08/2023	29/08/2023
									A071102	Failure to Interrogate	29/08/2023	29/08/2023
									A071204	No Pacing	29/08/2023	29/08/2023
					CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	29/08/2023	29/08/2023
									A071102	Failure to Interrogate	29/08/2023	29/08/2023
									A071204	No Pacing	29/08/2023	29/08/2023
							F1901	Additional Surgery	A070504	Premature Discharge of Battery	29/08/2023	29/08/2023
									A071102	Failure to Interrogate	29/08/2023	29/08/2023
									A071204	No Pacing	29/08/2023	29/08/2023
					CT233	Surgical	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	29/08/2023	29/08/2023
									A071102	Failure to Interrogate	29/08/2023	29/08/2023
									A071204	No Pacing	29/08/2023	29/08/2023
							F1901	Additional Surgery	A070504	Premature Discharge of Battery	29/08/2023	29/08/2023
									A071102	Failure to Interrogate	29/08/2023	29/08/2023
									A071204	No Pacing	29/08/2023	29/08/2023
					CT2406	Implantable	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	29/08/2023	29/08/2023
									A071102	Failure to Interrogate	29/08/2023	29/08/2023
									A071204	No Pacing	29/08/2023	29/08/2023
							F1901	Additional Surgery	A070504	Premature Discharge of Battery	29/08/2023	29/08/2023
									A071102	Failure to Interrogate	29/08/2023	29/08/2023
									A071204	No Pacing	29/08/2023	29/08/2023
					CT2407	Active implantable	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	29/08/2023	29/08/2023
									A071102	Failure to Interrogate	29/08/2023	29/08/2023
									A071204	No Pacing	29/08/2023	29/08/2023
							F1901	Additional Surgery	A070504	Premature Discharge of Battery	29/08/2023	29/08/2023
									A071102	Failure to Interrogate	29/08/2023	29/08/2023
									A071204	No Pacing	29/08/2023	29/08/2023
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	29/08/2023	29/08/2023
									A071102	Failure to Interrogate	29/08/2023	29/08/2023
									A071204	No Pacing	29/08/2023	29/08/2023
							F1901	Additional Surgery	A070504	Premature Discharge of Battery	29/08/2023	29/08/2023
									A071102	Failure to Interrogate	29/08/2023	29/08/2023
									A071204	No Pacing	29/08/2023	29/08/2023
					CT2727	Active implant devices/ancillaries	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	29/08/2023	29/08/2023
									A071102	Failure to Interrogate	29/08/2023	29/08/2023
									A071204	No Pacing	29/08/2023	29/08/2023
							F1901	Additional Surgery	A070504	Premature Discharge of Battery	29/08/2023	29/08/2023
									A071102	Failure to Interrogate	29/08/2023	29/08/2023
									A071204	No Pacing	29/08/2023	29/08/2023
					CT2750	Implantable cardiac electrical stimulators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	29/08/2023	29/08/2023
									A071102	Failure to Interrogate	29/08/2023	29/08/2023
									A071204	No Pacing	29/08/2023	29/08/2023
							F1901	Additional Surgery	A070504	Premature Discharge of Battery	29/08/2023	29/08/2023
									A071102	Failure to Interrogate	29/08/2023	29/08/2023
									A071204	No Pacing	29/08/2023	29/08/2023
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	29/08/2023	29/08/2023
									A071102	Failure to Interrogate	29/08/2023	29/08/2023
									A071204	No Pacing	29/08/2023	29/08/2023
							F1901	Additional Surgery	A070504	Premature Discharge of Battery	29/08/2023	29/08/2023
									A071102	Failure to Interrogate	29/08/2023	29/08/2023
									A071204	No Pacing	29/08/2023	29/08/2023
					CT2985	Active	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	29/08/2023	29/08/2023
									A071102	Failure to Interrogate	29/08/2023	29/08/2023
									A071204	No Pacing	29/08/2023	29/08/2023
							F1901	Additional Surgery	A070504	Premature Discharge of Battery	29/08/2023	29/08/2023
									A071102	Failure to Interrogate	29/08/2023	29/08/2023
									A071204	No Pacing	29/08/2023	29/08/2023
					CT3001	Implantable metal	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	29/08/2023	29/08/2023
									A071102	Failure to Interrogate	29/08/2023	29/08/2023
									A071204	No Pacing	29/08/2023	29/08/2023
							F1901	Additional Surgery	A070504	Premature Discharge of Battery	29/08/2023	29/08/2023
									A071102	Failure to Interrogate	29/08/2023	29/08/2023
									A071204	No Pacing	29/08/2023	29/08/2023
					CT3002	Implantable synthetic polymer	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	29/08/2023	29/08/2023
									A071102	Failure to Interrogate	29/08/2023	29/08/2023
									A071204	No Pacing	29/08/2023	29/08/2023
							F1901	Additional Surgery	A070504	Premature Discharge of Battery	29/08/2023	29/08/2023
									A071102	Failure to Interrogate	29/08/2023	29/08/2023
									A071204	No Pacing	29/08/2023	29/08/2023
					CT3006	Implantable inorganic material	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	29/08/2023	29/08/2023
									A071102	Failure to Interrogate	29/08/2023	29/08/2023
									A071204	No Pacing	29/08/2023	29/08/2023
							F1901	Additional Surgery	A070504	Premature Discharge of Battery	29/08/2023	29/08/2023
									A071102	Failure to Interrogate	29/08/2023	29/08/2023
									A071204	No Pacing	29/08/2023	29/08/2023
					CT314	Electrically powered	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	29/08/2023	29/08/2023
									A071102	Failure to Interrogate	29/08/2023	29/08/2023

									A071204	No Pacing	29/08/2023	29/08/2023
							F1901	Additional Surgery	A070504	Premature Discharge of Battery	29/08/2023	29/08/2023
									A071102	Failure to Interrogate	29/08/2023	29/08/2023
									A071204	No Pacing	29/08/2023	29/08/2023
					CT321	Long-term surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	29/08/2023	29/08/2023
									A071102	Failure to Interrogate	29/08/2023	29/08/2023
									A071204	No Pacing	29/08/2023	29/08/2023
							F1901	Additional Surgery	A070504	Premature Discharge of Battery	29/08/2023	29/08/2023
									A071102	Failure to Interrogate	29/08/2023	29/08/2023
									A071204	No Pacing	29/08/2023	29/08/2023
					CT449	Generators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	29/08/2023	29/08/2023
									A071102	Failure to Interrogate	29/08/2023	29/08/2023
									A071204	No Pacing	29/08/2023	29/08/2023
							F1901	Additional Surgery	A070504	Premature Discharge of Battery	29/08/2023	29/08/2023
									A071102	Failure to Interrogate	29/08/2023	29/08/2023
									A071204	No Pacing	29/08/2023	29/08/2023
					CT981	Single use	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	29/08/2023	29/08/2023
									A071102	Failure to Interrogate	29/08/2023	29/08/2023
									A071204	No Pacing	29/08/2023	29/08/2023
							F1901	Additional Surgery	A070504	Premature Discharge of Battery	29/08/2023	29/08/2023
									A071102	Failure to Interrogate	29/08/2023	29/08/2023
									A071204	No Pacing	29/08/2023	29/08/2023
					CT983	Surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	29/08/2023	29/08/2023
									A071102	Failure to Interrogate	29/08/2023	29/08/2023
									A071204	No Pacing	29/08/2023	29/08/2023
							F1901	Additional Surgery	A070504	Premature Discharge of Battery	29/08/2023	29/08/2023
									A071102	Failure to Interrogate	29/08/2023	29/08/2023
									A071204	No Pacing	29/08/2023	29/08/2023
14506	2023	All other reportable incidents			CT1684	Cardiac pacemakers and associated devices	E2104	Electric Shock	A070909	Oversensing	29/02/2024	29/02/2024
									A071302	Inappropriate/Inadequate Shock/Stimulation	29/02/2024	29/02/2024
									A090801	Signal Artifact/Noise	29/02/2024	29/02/2024
							F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A070909	Oversensing	29/02/2024	29/02/2024
									A071302	Inappropriate/Inadequate Shock/Stimulation	29/02/2024	29/02/2024
									A090801	Signal Artifact/Noise	29/02/2024	29/02/2024
14507	2023	All other reportable incidents	47267	Single-chamber implantable pacemaker, rate-responsive	CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
14508	2023		47265	Dual-chamber implantable pacemaker, rate-responsive	CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A25	No Apparent Adverse Event	23/11/2023	23/11/2023
							F26	No Health Consequences or Impact	A25	No Apparent Adverse Event	23/11/2023	23/11/2023
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A25	No Apparent Adverse Event	23/11/2023	23/11/2023
							F26	No Health Consequences or Impact	A25	No Apparent Adverse Event	23/11/2023	23/11/2023
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A25	No Apparent Adverse Event	23/11/2023	23/11/2023
							F26	No Health Consequences or Impact	A25	No Apparent Adverse Event	23/11/2023	23/11/2023
14509	2023	All other reportable incidents	47267	Single-chamber implantable pacemaker, rate-responsive	CT141	Electrodes/Leads and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070102	High Capture Threshold	14/06/2023	14/06/2023
									A0705	Battery Problem	14/06/2023	14/06/2023
									A070504	Premature Discharge of Battery	14/06/2023	14/06/2023
									A072201	High Impedance	14/06/2023	14/06/2023
							F08	Hospitalization or Prolonged Hospitalization	A070102	High Capture Threshold	14/06/2023	14/06/2023
									A0705	Battery Problem	14/06/2023	14/06/2023
									A070504	Premature Discharge of Battery	14/06/2023	14/06/2023
									A072201	High Impedance	14/06/2023	14/06/2023
							F1903	Device Explantation	A070102	High Capture Threshold	14/06/2023	14/06/2023
									A0705	Battery Problem	14/06/2023	14/06/2023
									A070504	Premature Discharge of Battery	14/06/2023	14/06/2023
									A072201	High Impedance	14/06/2023	14/06/2023
					CT145	Cardiovascular devices	E2403	No Clinical Signs, Symptoms or Conditions	A070102	High Capture Threshold	14/06/2023	14/06/2023
									A0705	Battery Problem	14/06/2023	14/06/2023
									A070504	Premature Discharge of Battery	14/06/2023	14/06/2023
									A072201	High Impedance	14/06/2023	14/06/2023
							F08	Hospitalization or Prolonged Hospitalization	A070102	High Capture Threshold	14/06/2023	14/06/2023
									A0705	Battery Problem	14/06/2023	14/06/2023
									A070504	Premature Discharge of Battery	14/06/2023	14/06/2023
									A072201	High Impedance	14/06/2023	14/06/2023
							F1903	Device Explantation	A070102	High Capture Threshold	14/06/2023	14/06/2023
									A0705	Battery Problem	14/06/2023	14/06/2023
									A070504	Premature Discharge of Battery	14/06/2023	14/06/2023
									A072201	High Impedance	14/06/2023	14/06/2023
					CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070102	High Capture Threshold	14/06/2023	14/06/2023
									A0705	Battery Problem	14/06/2023	14/06/2023
									A070504	Premature Discharge of Battery	14/06/2023	14/06/2023
									A072201	High Impedance	14/06/2023	14/06/2023
							F08	Hospitalization or Prolonged Hospitalization	A070102	High Capture Threshold	14/06/2023	14/06/2023
									A0705	Battery Problem	14/06/2023	14/06/2023
									A070504	Premature Discharge of Battery	14/06/2023	14/06/2023
									A072201	High Impedance	14/06/2023	14/06/2023
							F1903	Device Explantation	A070102	High Capture Threshold	14/06/2023	14/06/2023
									A0705	Battery Problem	14/06/2023	14/06/2023
									A070504	Premature Discharge of Battery	14/06/2023	14/06/2023
									A072201	High Impedance	14/06/2023	14/06/2023
					CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A070102	High Capture Threshold	14/06/2023	14/06/2023
									A0705	Battery Problem	14/06/2023	14/06/2023
									A070504	Premature Discharge of Battery	14/06/2023	14/06/2023
									A072201	High Impedance	14/06/2023	14/06/2023
							F08	Hospitalization or Prolonged Hospitalization	A070102	High Capture Threshold	14/06/2023	14/06/2023
									A0705	Battery Problem	14/06/2023	14/06/2023

									A070504	Premature Discharge of Battery	14/06/2023	14/06/2023
									A072201	High Impedance	14/06/2023	14/06/2023
								F1903	Device Explantation	A070102	High Capture Threshold	14/06/2023	14/06/2023
										A0705	Battery Problem	14/06/2023	14/06/2023
										A070504	Premature Discharge of Battery	14/06/2023	14/06/2023
										A072201	High Impedance	14/06/2023	14/06/2023
						CT233	Surgical	E2403	No Clinical Signs, Symptoms or Conditions	A070102	High Capture Threshold	14/06/2023	14/06/2023
										A0705	Battery Problem	14/06/2023	14/06/2023
										A070504	Premature Discharge of Battery	14/06/2023	14/06/2023
										A072201	High Impedance	14/06/2023	14/06/2023
								F08	Hospitalization or Prolonged Hospitalization	A070102	High Capture Threshold	14/06/2023	14/06/2023
										A0705	Battery Problem	14/06/2023	14/06/2023
										A070504	Premature Discharge of Battery	14/06/2023	14/06/2023
										A072201	High Impedance	14/06/2023	14/06/2023
								F1903	Device Explantation	A070102	High Capture Threshold	14/06/2023	14/06/2023
										A0705	Battery Problem	14/06/2023	14/06/2023
										A070504	Premature Discharge of Battery	14/06/2023	14/06/2023
										A072201	High Impedance	14/06/2023	14/06/2023
						CT2406	Implantable	E2403	No Clinical Signs, Symptoms or Conditions	A070102	High Capture Threshold	14/06/2023	14/06/2023
										A0705	Battery Problem	14/06/2023	14/06/2023
										A070504	Premature Discharge of Battery	14/06/2023	14/06/2023
										A072201	High Impedance	14/06/2023	14/06/2023
								F08	Hospitalization or Prolonged Hospitalization	A070102	High Capture Threshold	14/06/2023	14/06/2023
										A0705	Battery Problem	14/06/2023	14/06/2023
										A070504	Premature Discharge of Battery	14/06/2023	14/06/2023
										A072201	High Impedance	14/06/2023	14/06/2023
								F1903	Device Explantation	A070102	High Capture Threshold	14/06/2023	14/06/2023
										A0705	Battery Problem	14/06/2023	14/06/2023
										A070504	Premature Discharge of Battery	14/06/2023	14/06/2023
										A072201	High Impedance	14/06/2023	14/06/2023
						CT2407	Active implantable	E2403	No Clinical Signs, Symptoms or Conditions	A070102	High Capture Threshold	14/06/2023	14/06/2023
										A0705	Battery Problem	14/06/2023	14/06/2023
										A070504	Premature Discharge of Battery	14/06/2023	14/06/2023
										A072201	High Impedance	14/06/2023	14/06/2023
								F08	Hospitalization or Prolonged Hospitalization	A070102	High Capture Threshold	14/06/2023	14/06/2023
										A0705	Battery Problem	14/06/2023	14/06/2023
										A070504	Premature Discharge of Battery	14/06/2023	14/06/2023
										A072201	High Impedance	14/06/2023	14/06/2023
								F1903	Device Explantation	A070102	High Capture Threshold	14/06/2023	14/06/2023
										A0705	Battery Problem	14/06/2023	14/06/2023
										A070504	Premature Discharge of Battery	14/06/2023	14/06/2023
										A072201	High Impedance	14/06/2023	14/06/2023
						CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A070102	High Capture Threshold	14/06/2023	14/06/2023
										A0705	Battery Problem	14/06/2023	14/06/2023
										A070504	Premature Discharge of Battery	14/06/2023	14/06/2023
										A072201	High Impedance	14/06/2023	14/06/2023
								F08	Hospitalization or Prolonged Hospitalization	A070102	High Capture Threshold	14/06/2023	14/06/2023
										A0705	Battery Problem	14/06/2023	14/06/2023
										A070504	Premature Discharge of Battery	14/06/2023	14/06/2023
										A072201	High Impedance	14/06/2023	14/06/2023
								F1903	Device Explantation	A070102	High Capture Threshold	14/06/2023	14/06/2023
										A0705	Battery Problem	14/06/2023	14/06/2023
										A070504	Premature Discharge of Battery	14/06/2023	14/06/2023
										A072201	High Impedance	14/06/2023	14/06/2023
						CT2727	Active implant devices/ancillaries	E2403	No Clinical Signs, Symptoms or Conditions	A070102	High Capture Threshold	14/06/2023	14/06/2023
										A0705	Battery Problem	14/06/2023	14/06/2023
										A070504	Premature Discharge of Battery	14/06/2023	14/06/2023
										A072201	High Impedance	14/06/2023	14/06/2023
								F08	Hospitalization or Prolonged Hospitalization	A070102	High Capture Threshold	14/06/2023	14/06/2023
										A0705	Battery Problem	14/06/2023	14/06/2023
										A070504	Premature Discharge of Battery	14/06/2023	14/06/2023
										A072201	High Impedance	14/06/2023	14/06/2023
								F1903	Device Explantation	A070102	High Capture Threshold	14/06/2023	14/06/2023
										A0705	Battery Problem	14/06/2023	14/06/2023
										A070504	Premature Discharge of Battery	14/06/2023	14/06/2023
										A072201	High Impedance	14/06/2023	14/06/2023
						CT2750	Implantable cardiac electrical stimulators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070102	High Capture Threshold	14/06/2023	14/06/2023
										A0705	Battery Problem	14/06/2023	14/06/2023
										A070504	Premature Discharge of Battery	14/06/2023	14/06/2023
										A072201	High Impedance	14/06/2023	14/06/2023
								F08	Hospitalization or Prolonged Hospitalization	A070102	High Capture Threshold	14/06/2023	14/06/2023
										A0705	Battery Problem	14/06/2023	14/06/2023
										A070504	Premature Discharge of Battery	14/06/2023	14/06/2023
										A072201	High Impedance	14/06/2023	14/06/2023
								F1903	Device Explantation	A070102	High Capture Threshold	14/06/2023	14/06/2023
										A0705	Battery Problem	14/06/2023	14/06/2023
										A070504	Premature Discharge of Battery	14/06/2023	14/06/2023
										A072201	High Impedance	14/06/2023	14/06/2023
						CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A070102	High Capture Threshold	14/06/2023	14/06/2023
										A0705	Battery Problem	14/06/2023	14/06/2023
										A070504	Premature Discharge of Battery	14/06/2023	14/06/2023
										A072201	High Impedance	14/06/2023	14/06/2023
								F08	Hospitalization or Prolonged Hospitalization	A070102	High Capture Threshold	14/06/2023	14/06/2023
										A0705	Battery Problem	14/06/2023	14/06/2023
										A070504	Premature Discharge of Battery	14/06/2023	14/06/2023
										A072201	High Impedance	14/06/2023	14/06/2023
								F1903	Device Explantation	A070102	High Capture Threshold	14/06/2023	14/06/2023
										A0705	Battery Problem	14/06/2023	14/06/2023

								A070504	Premature Discharge of Battery	14/06/2023	14/06/2023
								A072201	High Impedance	14/06/2023	14/06/2023
					CT2985	Active	E2403	No Clinical Signs, Symptoms or Conditions	A070102	High Capture Threshold	14/06/2023	14/06/2023
									A0705	Battery Problem	14/06/2023	14/06/2023
									A070504	Premature Discharge of Battery	14/06/2023	14/06/2023
									A072201	High Impedance	14/06/2023	14/06/2023
							F08	Hospitalization or Prolonged Hospitalization	A070102	High Capture Threshold	14/06/2023	14/06/2023
									A0705	Battery Problem	14/06/2023	14/06/2023
									A070504	Premature Discharge of Battery	14/06/2023	14/06/2023
									A072201	High Impedance	14/06/2023	14/06/2023
							F1903	Device Explantation	A070102	High Capture Threshold	14/06/2023	14/06/2023
									A0705	Battery Problem	14/06/2023	14/06/2023
									A070504	Premature Discharge of Battery	14/06/2023	14/06/2023
									A072201	High Impedance	14/06/2023	14/06/2023
					CT3001	Implantable metal	E2403	No Clinical Signs, Symptoms or Conditions	A070102	High Capture Threshold	14/06/2023	14/06/2023
									A0705	Battery Problem	14/06/2023	14/06/2023
									A070504	Premature Discharge of Battery	14/06/2023	14/06/2023
									A072201	High Impedance	14/06/2023	14/06/2023
							F08	Hospitalization or Prolonged Hospitalization	A070102	High Capture Threshold	14/06/2023	14/06/2023
									A0705	Battery Problem	14/06/2023	14/06/2023
									A070504	Premature Discharge of Battery	14/06/2023	14/06/2023
									A072201	High Impedance	14/06/2023	14/06/2023
							F1903	Device Explantation	A070102	High Capture Threshold	14/06/2023	14/06/2023
									A0705	Battery Problem	14/06/2023	14/06/2023
									A070504	Premature Discharge of Battery	14/06/2023	14/06/2023
									A072201	High Impedance	14/06/2023	14/06/2023
					CT3002	Implantable synthetic polymer	E2403	No Clinical Signs, Symptoms or Conditions	A070102	High Capture Threshold	14/06/2023	14/06/2023
									A0705	Battery Problem	14/06/2023	14/06/2023
									A070504	Premature Discharge of Battery	14/06/2023	14/06/2023
									A072201	High Impedance	14/06/2023	14/06/2023
							F08	Hospitalization or Prolonged Hospitalization	A070102	High Capture Threshold	14/06/2023	14/06/2023
									A0705	Battery Problem	14/06/2023	14/06/2023
									A070504	Premature Discharge of Battery	14/06/2023	14/06/2023
									A072201	High Impedance	14/06/2023	14/06/2023
							F1903	Device Explantation	A070102	High Capture Threshold	14/06/2023	14/06/2023
									A0705	Battery Problem	14/06/2023	14/06/2023
									A070504	Premature Discharge of Battery	14/06/2023	14/06/2023
									A072201	High Impedance	14/06/2023	14/06/2023
					CT3006	Implantable inorganic material	E2403	No Clinical Signs, Symptoms or Conditions	A070102	High Capture Threshold	14/06/2023	14/06/2023
									A0705	Battery Problem	14/06/2023	14/06/2023
									A070504	Premature Discharge of Battery	14/06/2023	14/06/2023
									A072201	High Impedance	14/06/2023	14/06/2023
							F08	Hospitalization or Prolonged Hospitalization	A070102	High Capture Threshold	14/06/2023	14/06/2023
									A0705	Battery Problem	14/06/2023	14/06/2023
									A070504	Premature Discharge of Battery	14/06/2023	14/06/2023
									A072201	High Impedance	14/06/2023	14/06/2023
							F1903	Device Explantation	A070102	High Capture Threshold	14/06/2023	14/06/2023
									A0705	Battery Problem	14/06/2023	14/06/2023
									A070504	Premature Discharge of Battery	14/06/2023	14/06/2023
									A072201	High Impedance	14/06/2023	14/06/2023
					CT314	Electrically powered	E2403	No Clinical Signs, Symptoms or Conditions	A070102	High Capture Threshold	14/06/2023	14/06/2023
									A0705	Battery Problem	14/06/2023	14/06/2023
									A070504	Premature Discharge of Battery	14/06/2023	14/06/2023
									A072201	High Impedance	14/06/2023	14/06/2023
							F08	Hospitalization or Prolonged Hospitalization	A070102	High Capture Threshold	14/06/2023	14/06/2023
									A0705	Battery Problem	14/06/2023	14/06/2023
									A070504	Premature Discharge of Battery	14/06/2023	14/06/2023
									A072201	High Impedance	14/06/2023	14/06/2023
							F1903	Device Explantation	A070102	High Capture Threshold	14/06/2023	14/06/2023
									A0705	Battery Problem	14/06/2023	14/06/2023
									A070504	Premature Discharge of Battery	14/06/2023	14/06/2023
									A072201	High Impedance	14/06/2023	14/06/2023
					CT321	Long-term surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A070102	High Capture Threshold	14/06/2023	14/06/2023
									A0705	Battery Problem	14/06/2023	14/06/2023
									A070504	Premature Discharge of Battery	14/06/2023	14/06/2023
									A072201	High Impedance	14/06/2023	14/06/2023
							F08	Hospitalization or Prolonged Hospitalization	A070102	High Capture Threshold	14/06/2023	14/06/2023
									A0705	Battery Problem	14/06/2023	14/06/2023
									A070504	Premature Discharge of Battery	14/06/2023	14/06/2023
									A072201	High Impedance	14/06/2023	14/06/2023
							F1903	Device Explantation	A070102	High Capture Threshold	14/06/2023	14/06/2023
									A0705	Battery Problem	14/06/2023	14/06/2023
									A070504	Premature Discharge of Battery	14/06/2023	14/06/2023
									A072201	High Impedance	14/06/2023	14/06/2023
					CT449	Generators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070102	High Capture Threshold	14/06/2023	14/06/2023
									A0705	Battery Problem	14/06/2023	14/06/2023
									A070504	Premature Discharge of Battery	14/06/2023	14/06/2023
									A072201	High Impedance	14/06/2023	14/06/2023
							F08	Hospitalization or Prolonged Hospitalization	A070102	High Capture Threshold	14/06/2023	14/06/2023
									A0705	Battery Problem	14/06/2023	14/06/2023
									A070504	Premature Discharge of Battery	14/06/2023	14/06/2023
									A072201	High Impedance	14/06/2023	14/06/2023
							F1903	Device Explantation	A070102	High Capture Threshold	14/06/2023	14/06/2023
									A0705	Battery Problem	14/06/2023	14/06/2023
									A070504	Premature Discharge of Battery	14/06/2023	14/06/2023
									A072201	High Impedance	14/06/2023	14/06/2023
					CT981	Single use	E2403	No Clinical Signs, Symptoms or Conditions	A070102	High Capture Threshold	14/06/2023	14/06/2023
									A0705	Battery Problem	14/06/2023	14/06/2023

								A070504	Premature Discharge of Battery	14/06/2023	14/06/2023
								A072201	High Impedance	14/06/2023	14/06/2023
						F08	Hospitalization or Prolonged Hospitalization	A070102	High Capture Threshold	14/06/2023	14/06/2023
								A0705	Battery Problem	14/06/2023	14/06/2023
								A070504	Premature Discharge of Battery	14/06/2023	14/06/2023
								A072201	High Impedance	14/06/2023	14/06/2023
						F1903	Device Explantation	A070102	High Capture Threshold	14/06/2023	14/06/2023
								A0705	Battery Problem	14/06/2023	14/06/2023
								A070504	Premature Discharge of Battery	14/06/2023	14/06/2023
								A072201	High Impedance	14/06/2023	14/06/2023
				CT983	Surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A070102	High Capture Threshold	14/06/2023	14/06/2023
								A0705	Battery Problem	14/06/2023	14/06/2023
								A070504	Premature Discharge of Battery	14/06/2023	14/06/2023
								A072201	High Impedance	14/06/2023	14/06/2023
						F08	Hospitalization or Prolonged Hospitalization	A070102	High Capture Threshold	14/06/2023	14/06/2023
								A0705	Battery Problem	14/06/2023	14/06/2023
								A070504	Premature Discharge of Battery	14/06/2023	14/06/2023
								A072201	High Impedance	14/06/2023	14/06/2023
						F1903	Device Explantation	A070102	High Capture Threshold	14/06/2023	14/06/2023
								A0705	Battery Problem	14/06/2023	14/06/2023
								A070504	Premature Discharge of Battery	14/06/2023	14/06/2023
								A072201	High Impedance	14/06/2023	14/06/2023
14510	2023	All other reportable incidents	35852	Single-chamber implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A0709	Device Sensing Problem		
									A070909	Oversensing		
							F26	No Health Consequences or Impact	A0709	Device Sensing Problem		
									A070909	Oversensing		
14511	2023		47267	Single-chamber implantable pacemaker, rate-responsive	CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	16/11/2023	16/11/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	16/11/2023	16/11/2023
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	16/11/2023	16/11/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	16/11/2023	16/11/2023
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	16/11/2023	16/11/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	16/11/2023	16/11/2023
14512	2023	All other reportable incidents	37265	Dual-chamber implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A070909	Oversensing		
							F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A070909	Oversensing		
14513	2023	Death	37265	Dual-chamber implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E2401	Insufficient Information	A071201	Failure to Convert Rhythm		
									A072201	High Impedance		
									A0908	Incorrect, Inadequate or Imprecise Result or Readings		
							F02	Death	A071201	Failure to Convert Rhythm		
									A072201	High Impedance		
									A0908	Incorrect, Inadequate or Imprecise Result or Readings		
14514	2023	All other reportable incidents	17882	Physiologic-monitoring defibrillation system	CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A08	Calibration Problem		
							F27	Problem Identified During Non-Clinical Procedure	A08	Calibration Problem		
14515	2023	Unanticipated serious deterioration in state of health	47205	Cardiac pulse generator programmer	CT1684	Cardiac pacemakers and associated devices	E060101	Asystole	A0902	Display or Visual Feedback Problem		
									A1104	Computer Operating System Problem		
									A13	Communication or Transmission Problem		
							F1203	Life Threatening Illness or Injury	A0902	Display or Visual Feedback Problem		
									A1104	Computer Operating System Problem		
									A13	Communication or Transmission Problem		
							F2306	Resuscitation	A0902	Display or Visual Feedback Problem		
									A1104	Computer Operating System Problem		
									A13	Communication or Transmission Problem		
14516	2023	All other reportable incidents	47265	Dual-chamber implantable pacemaker, rate-responsive	CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
14517	2023	All other reportable incidents	17882	Physiologic-monitoring defibrillation system	CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A071204	No Pacing		
							F27	Problem Identified During Non-Clinical Procedure	A071204	No Pacing		
14518	2023		47265	Dual-chamber implantable pacemaker, rate-responsive	CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	17/11/2023	17/11/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	17/11/2023	17/11/2023
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	17/11/2023	17/11/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	17/11/2023	17/11/2023
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	17/11/2023	17/11/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	17/11/2023	17/11/2023
14519	2023	All other reportable incidents	47267	Single-chamber implantable pacemaker, rate-responsive	CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem	19/06/2023	19/06/2023
									A070504	Premature Discharge of Battery	19/06/2023	19/06/2023
							F08	Hospitalization or Prolonged Hospitalization	A0705	Battery Problem	19/06/2023	19/06/2023
									A070504	Premature Discharge of Battery	19/06/2023	19/06/2023
							F1905	Device Revision or Replacement	A0705	Battery Problem	19/06/2023	19/06/2023
									A070504	Premature Discharge of Battery	19/06/2023	19/06/2023
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem	19/06/2023	19/06/2023
									A070504	Premature Discharge of Battery	19/06/2023	19/06/2023
							F08	Hospitalization or Prolonged Hospitalization	A0705	Battery Problem	19/06/2023	19/06/2023
									A070504	Premature Discharge of Battery	19/06/2023	19/06/2023
							F1905	Device Revision or Replacement	A0705	Battery Problem	19/06/2023	19/06/2023
									A070504	Premature Discharge of Battery	19/06/2023	19/06/2023
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem	19/06/2023	19/06/2023
									A070504	Premature Discharge of Battery	19/06/2023	19/06/2023
							F08	Hospitalization or Prolonged Hospitalization	A0705	Battery Problem	19/06/2023	19/06/2023

								A070504	Premature Discharge of Battery	19/06/2023	19/06/2023
						F1905	Device Revision or Replacement	A0705	Battery Problem	19/06/2023	19/06/2023
								A070504	Premature Discharge of Battery	19/06/2023	19/06/2023
14520	2023	All other reportable incidents	47265	Dual-chamber implantable pacemaker, rate-responsive	CT141	Electrodes/Leads and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT145	Cardiovascular devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT233	Surgical	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT2406	Implantable	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT2407	Active implantable	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT2727	Active implant devices/ancillaries	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT2750	Implantable cardiac electrical stimulators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT2985	Active	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT3001	Implantable metal	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT3002	Implantable synthetic polymer	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT3006	Implantable inorganic material	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT314	Electrically powered	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT321	Long-term surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT449	Generators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT981	Single use	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT983	Surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
14521	2023	All other reportable incidents	17882	Physiologic-monitoring defibrillation system	CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A0713	Defibrillation/Stimulation Problem		
							F27	Problem Identified During Non-Clinical Procedure	A0713	Defibrillation/Stimulation Problem		
14522	2023		47265	Dual-chamber implantable pacemaker, rate-responsive	CT141	Electrodes/Leads and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem		
									A070503	Failure to Run on Battery		
							F26	No Health Consequences or Impact	A0705	Battery Problem		
									A070503	Failure to Run on Battery		
					CT145	Cardiovascular devices	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem		
									A070503	Failure to Run on Battery		
							F26	No Health Consequences or Impact	A0705	Battery Problem		
									A070503	Failure to Run on Battery		
					CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem		
									A070503	Failure to Run on Battery		
							F26	No Health Consequences or Impact	A0705	Battery Problem		
									A070503	Failure to Run on Battery		
					CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem		
									A070503	Failure to Run on Battery		
							F26	No Health Consequences or Impact	A0705	Battery Problem		
									A070503	Failure to Run on Battery		
					CT233	Surgical	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem		
									A070503	Failure to Run on Battery		
							F26	No Health Consequences or Impact	A0705	Battery Problem		
									A070503	Failure to Run on Battery		
					CT2406	Implantable	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem		
									A070503	Failure to Run on Battery		
							F26	No Health Consequences or Impact	A0705	Battery Problem		
									A070503	Failure to Run on Battery		
					CT2407	Active implantable	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem		
									A070503	Failure to Run on Battery		
							F26	No Health Consequences or Impact	A0705	Battery Problem		
									A070503	Failure to Run on Battery		
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem		
									A070503	Failure to Run on Battery		
							F26	No Health Consequences or Impact	A0705	Battery Problem		
									A070503	Failure to Run on Battery		
					CT2727	Active implant devices/ancillaries	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem		
									A070503	Failure to Run on Battery		
							F26	No Health Consequences or Impact	A0705	Battery Problem		
									A070503	Failure to Run on Battery		
					CT2750	Implantable cardiac electrical stimulators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem		
									A070503	Failure to Run on Battery		
							F26	No Health Consequences or Impact	A0705	Battery Problem		
									A070503	Failure to Run on Battery		
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem		
									A070503	Failure to Run on Battery		
							F26	No Health Consequences or Impact	A0705	Battery Problem		

									A070503	Failure to Run on Battery		
					CT2985	Active	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem		
									A070503	Failure to Run on Battery		
							F26	No Health Consequences or Impact	A0705	Battery Problem		
									A070503	Failure to Run on Battery		
					CT3001	Implantable metal	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem		
									A070503	Failure to Run on Battery		
							F26	No Health Consequences or Impact	A0705	Battery Problem		
									A070503	Failure to Run on Battery		
					CT3002	Implantable synthetic polymer	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem		
									A070503	Failure to Run on Battery		
							F26	No Health Consequences or Impact	A0705	Battery Problem		
									A070503	Failure to Run on Battery		
					CT3006	Implantable inorganic material	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem		
									A070503	Failure to Run on Battery		
							F26	No Health Consequences or Impact	A0705	Battery Problem		
									A070503	Failure to Run on Battery		
					CT314	Electrically powered	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem		
									A070503	Failure to Run on Battery		
							F26	No Health Consequences or Impact	A0705	Battery Problem		
									A070503	Failure to Run on Battery		
					CT321	Long-term surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem		
									A070503	Failure to Run on Battery		
							F26	No Health Consequences or Impact	A0705	Battery Problem		
									A070503	Failure to Run on Battery		
					CT449	Generators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem		
									A070503	Failure to Run on Battery		
							F26	No Health Consequences or Impact	A0705	Battery Problem		
									A070503	Failure to Run on Battery		
					CT981	Single use	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem		
									A070503	Failure to Run on Battery		
							F26	No Health Consequences or Impact	A0705	Battery Problem		
									A070503	Failure to Run on Battery		
					CT983	Surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem		
									A070503	Failure to Run on Battery		
							F26	No Health Consequences or Impact	A0705	Battery Problem		
									A070503	Failure to Run on Battery		
14523	2023		47265	Dual-chamber implantable pacemaker, rate-responsive	CT141	Electrodes/Leads and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	24/11/2023	24/11/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	24/11/2023	24/11/2023
					CT145	Cardiovascular devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	24/11/2023	24/11/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	24/11/2023	24/11/2023
					CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	24/11/2023	24/11/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	24/11/2023	24/11/2023
					CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	24/11/2023	24/11/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	24/11/2023	24/11/2023
					CT233	Surgical	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	24/11/2023	24/11/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	24/11/2023	24/11/2023
					CT2406	Implantable	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	24/11/2023	24/11/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	24/11/2023	24/11/2023
					CT2407	Active implantable	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	24/11/2023	24/11/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	24/11/2023	24/11/2023
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	24/11/2023	24/11/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	24/11/2023	24/11/2023
					CT2727	Active implant devices/ancillaries	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	24/11/2023	24/11/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	24/11/2023	24/11/2023
					CT2750	Implantable cardiac electrical stimulators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	24/11/2023	24/11/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	24/11/2023	24/11/2023
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	24/11/2023	24/11/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	24/11/2023	24/11/2023
					CT2985	Active	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	24/11/2023	24/11/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	24/11/2023	24/11/2023
					CT3001	Implantable metal	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	24/11/2023	24/11/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	24/11/2023	24/11/2023
					CT3002	Implantable synthetic polymer	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	24/11/2023	24/11/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	24/11/2023	24/11/2023
					CT3006	Implantable inorganic material	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	24/11/2023	24/11/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	24/11/2023	24/11/2023
					CT314	Electrically powered	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	24/11/2023	24/11/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	24/11/2023	24/11/2023
					CT321	Long-term surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	24/11/2023	24/11/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	24/11/2023	24/11/2023
					CT449	Generators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	24/11/2023	24/11/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	24/11/2023	24/11/2023
					CT981	Single use	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	24/11/2023	24/11/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	24/11/2023	24/11/2023
					CT983	Surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	24/11/2023	24/11/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	24/11/2023	24/11/2023
14524	2023		47265	Dual-chamber implantable pacemaker, rate-responsive	CT141	Electrodes/Leads and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	24/11/2023	24/11/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	24/11/2023	24/11/2023
							F1903	Device Explantation	A24	No Apparent Device Problem	24/11/2023	24/11/2023
							F1905	Device Revision or Replacement	A24	No Apparent Device Problem	24/11/2023	24/11/2023
					CT145	Cardiovascular devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	24/11/2023	24/11/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	24/11/2023	24/11/2023
							F1903	Device Explantation	A24	No Apparent Device Problem	24/11/2023	24/11/2023
							F1905	Device Revision or Replacement	A24	No Apparent Device Problem	24/11/2023	24/11/2023
					CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	24/11/2023	24/11/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	24/11/2023	24/11/2023
							F1903	Device Explantation	A24	No Apparent Device Problem	24/11/2023	24/11/2023

							F1905	Device Revision or Replacement	A24	No Apparent Device Problem	24/11/2023	24/11/2023
					CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	24/11/2023	24/11/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	24/11/2023	24/11/2023
							F1903	Device Explantation	A24	No Apparent Device Problem	24/11/2023	24/11/2023
							F1905	Device Revision or Replacement	A24	No Apparent Device Problem	24/11/2023	24/11/2023
					CT233	Surgical	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	24/11/2023	24/11/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	24/11/2023	24/11/2023
							F1903	Device Explantation	A24	No Apparent Device Problem	24/11/2023	24/11/2023
							F1905	Device Revision or Replacement	A24	No Apparent Device Problem	24/11/2023	24/11/2023
					CT2406	Implantable	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	24/11/2023	24/11/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	24/11/2023	24/11/2023
							F1903	Device Explantation	A24	No Apparent Device Problem	24/11/2023	24/11/2023
							F1905	Device Revision or Replacement	A24	No Apparent Device Problem	24/11/2023	24/11/2023
					CT2407	Active implantable	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	24/11/2023	24/11/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	24/11/2023	24/11/2023
							F1903	Device Explantation	A24	No Apparent Device Problem	24/11/2023	24/11/2023
							F1905	Device Revision or Replacement	A24	No Apparent Device Problem	24/11/2023	24/11/2023
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	24/11/2023	24/11/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	24/11/2023	24/11/2023
							F1903	Device Explantation	A24	No Apparent Device Problem	24/11/2023	24/11/2023
							F1905	Device Revision or Replacement	A24	No Apparent Device Problem	24/11/2023	24/11/2023
					CT2727	Active implant devices/ancillaries	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	24/11/2023	24/11/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	24/11/2023	24/11/2023
							F1903	Device Explantation	A24	No Apparent Device Problem	24/11/2023	24/11/2023
							F1905	Device Revision or Replacement	A24	No Apparent Device Problem	24/11/2023	24/11/2023
					CT2750	Implantable cardiac electrical stimulators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	24/11/2023	24/11/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	24/11/2023	24/11/2023
							F1903	Device Explantation	A24	No Apparent Device Problem	24/11/2023	24/11/2023
							F1905	Device Revision or Replacement	A24	No Apparent Device Problem	24/11/2023	24/11/2023
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	24/11/2023	24/11/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	24/11/2023	24/11/2023
							F1903	Device Explantation	A24	No Apparent Device Problem	24/11/2023	24/11/2023
							F1905	Device Revision or Replacement	A24	No Apparent Device Problem	24/11/2023	24/11/2023
					CT2985	Active	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	24/11/2023	24/11/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	24/11/2023	24/11/2023
							F1903	Device Explantation	A24	No Apparent Device Problem	24/11/2023	24/11/2023
							F1905	Device Revision or Replacement	A24	No Apparent Device Problem	24/11/2023	24/11/2023
					CT3001	Implantable metal	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	24/11/2023	24/11/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	24/11/2023	24/11/2023
							F1903	Device Explantation	A24	No Apparent Device Problem	24/11/2023	24/11/2023
							F1905	Device Revision or Replacement	A24	No Apparent Device Problem	24/11/2023	24/11/2023
					CT3002	Implantable synthetic polymer	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	24/11/2023	24/11/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	24/11/2023	24/11/2023
							F1903	Device Explantation	A24	No Apparent Device Problem	24/11/2023	24/11/2023
							F1905	Device Revision or Replacement	A24	No Apparent Device Problem	24/11/2023	24/11/2023
					CT3006	Implantable inorganic material	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	24/11/2023	24/11/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	24/11/2023	24/11/2023
							F1903	Device Explantation	A24	No Apparent Device Problem	24/11/2023	24/11/2023
							F1905	Device Revision or Replacement	A24	No Apparent Device Problem	24/11/2023	24/11/2023
					CT314	Electrically powered	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	24/11/2023	24/11/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	24/11/2023	24/11/2023
							F1903	Device Explantation	A24	No Apparent Device Problem	24/11/2023	24/11/2023
							F1905	Device Revision or Replacement	A24	No Apparent Device Problem	24/11/2023	24/11/2023
					CT321	Long-term surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	24/11/2023	24/11/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	24/11/2023	24/11/2023
							F1903	Device Explantation	A24	No Apparent Device Problem	24/11/2023	24/11/2023
							F1905	Device Revision or Replacement	A24	No Apparent Device Problem	24/11/2023	24/11/2023
					CT449	Generators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	24/11/2023	24/11/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	24/11/2023	24/11/2023
							F1903	Device Explantation	A24	No Apparent Device Problem	24/11/2023	24/11/2023
							F1905	Device Revision or Replacement	A24	No Apparent Device Problem	24/11/2023	24/11/2023
					CT981	Single use	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	24/11/2023	24/11/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	24/11/2023	24/11/2023
							F1903	Device Explantation	A24	No Apparent Device Problem	24/11/2023	24/11/2023
							F1905	Device Revision or Replacement	A24	No Apparent Device Problem	24/11/2023	24/11/2023
					CT983	Surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	24/11/2023	24/11/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	24/11/2023	24/11/2023
							F1903	Device Explantation	A24	No Apparent Device Problem	24/11/2023	24/11/2023
							F1905	Device Revision or Replacement	A24	No Apparent Device Problem	24/11/2023	24/11/2023
14525	2023		47265	Dual-chamber implantable pacemaker, rate-responsive	CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	24/11/2023	24/11/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	24/11/2023	24/11/2023
							F1903	Device Explantation	A24	No Apparent Device Problem	24/11/2023	24/11/2023
							F1905	Device Revision or Replacement	A24	No Apparent Device Problem	24/11/2023	24/11/2023
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	24/11/2023	24/11/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	24/11/2023	24/11/2023
							F1903	Device Explantation	A24	No Apparent Device Problem	24/11/2023	24/11/2023
							F1905	Device Revision or Replacement	A24	No Apparent Device Problem	24/11/2023	24/11/2023
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	24/11/2023	24/11/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	24/11/2023	24/11/2023
							F1903	Device Explantation	A24	No Apparent Device Problem	24/11/2023	24/11/2023
							F1905	Device Revision or Replacement	A24	No Apparent Device Problem	24/11/2023	24/11/2023
14526	2023	All other reportable incidents	63188	Left-ventricular wireless pacing system sensor/transmitter	CT141	Electrodes/Leads and associated devices			A070801	Complete Loss of Power		
					CT145	Cardiovascular devices			A070801	Complete Loss of Power		
					CT1577	Wireless monitoring system transmitters			A070801	Complete Loss of Power		
					CT1684	Cardiac pacemakers and associated devices			A070801	Complete Loss of Power		
					CT2317	Transmitters/Receivers/Transceivers			A070801	Complete Loss of Power		
					CT233	Surgical			A070801	Complete Loss of Power		
					CT2406	Implantable			A070801	Complete Loss of Power		

					CT2407	Active implantable			A070801	Complete Loss of Power		
					CT255	Transducers and associated devices			A070801	Complete Loss of Power		
					CT2726	Physiological monitoring/measuring devices			A070801	Complete Loss of Power		
					CT2727	Active implant devices/ancillaries			A070801	Complete Loss of Power		
					CT2750	Implantable cardiac electrical stimulators and associated devices			A070801	Complete Loss of Power		
					CT2885	Physiological monitoring/measuring system transmitters/receivers/transceivers			A070801	Complete Loss of Power		
					CT2985	Active			A070801	Complete Loss of Power		
					CT3001	Implantable metal			A070801	Complete Loss of Power		
					CT3002	Implantable synthetic polymer			A070801	Complete Loss of Power		
					CT3006	Implantable inorganic material			A070801	Complete Loss of Power		
					CT314	Electrically powered			A070801	Complete Loss of Power		
					CT321	Long-term surgical invasive			A070801	Complete Loss of Power		
					CT384	Sensors and associated devices			A070801	Complete Loss of Power		
					CT449	Generators and associated devices			A070801	Complete Loss of Power		
					CT981	Single use			A070801	Complete Loss of Power		
					CT983	Surgical invasive			A070801	Complete Loss of Power		
					CT994	Sensors			A070801	Complete Loss of Power		
14527	2023	All other reportable incidents	47265	Dual-chamber implantable pacemaker, rate-responsive	CT141	Electrodes/Leads and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A0908	Incorrect, Inadequate or Imprecise Result or Readings	08/08/2023	08/08/2023
							F08	Hospitalization or Prolonged Hospitalization	A0908	Incorrect, Inadequate or Imprecise Result or Readings	08/08/2023	08/08/2023
							F1905	Device Revision or Replacement	A0908	Incorrect, Inadequate or Imprecise Result or Readings	08/08/2023	08/08/2023
					CT145	Cardiovascular devices	E2403	No Clinical Signs, Symptoms or Conditions	A0908	Incorrect, Inadequate or Imprecise Result or Readings	08/08/2023	08/08/2023
							F08	Hospitalization or Prolonged Hospitalization	A0908	Incorrect, Inadequate or Imprecise Result or Readings	08/08/2023	08/08/2023
							F1905	Device Revision or Replacement	A0908	Incorrect, Inadequate or Imprecise Result or Readings	08/08/2023	08/08/2023
					CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A0908	Incorrect, Inadequate or Imprecise Result or Readings	08/08/2023	08/08/2023
							F08	Hospitalization or Prolonged Hospitalization	A0908	Incorrect, Inadequate or Imprecise Result or Readings	08/08/2023	08/08/2023
							F1905	Device Revision or Replacement	A0908	Incorrect, Inadequate or Imprecise Result or Readings	08/08/2023	08/08/2023
					CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A0908	Incorrect, Inadequate or Imprecise Result or Readings	08/08/2023	08/08/2023
							F08	Hospitalization or Prolonged Hospitalization	A0908	Incorrect, Inadequate or Imprecise Result or Readings	08/08/2023	08/08/2023
							F1905	Device Revision or Replacement	A0908	Incorrect, Inadequate or Imprecise Result or Readings	08/08/2023	08/08/2023
					CT233	Surgical	E2403	No Clinical Signs, Symptoms or Conditions	A0908	Incorrect, Inadequate or Imprecise Result or Readings	08/08/2023	08/08/2023
							F08	Hospitalization or Prolonged Hospitalization	A0908	Incorrect, Inadequate or Imprecise Result or Readings	08/08/2023	08/08/2023
							F1905	Device Revision or Replacement	A0908	Incorrect, Inadequate or Imprecise Result or Readings	08/08/2023	08/08/2023
					CT2406	Implantable	E2403	No Clinical Signs, Symptoms or Conditions	A0908	Incorrect, Inadequate or Imprecise Result or Readings	08/08/2023	08/08/2023
							F08	Hospitalization or Prolonged Hospitalization	A0908	Incorrect, Inadequate or Imprecise Result or Readings	08/08/2023	08/08/2023
							F1905	Device Revision or Replacement	A0908	Incorrect, Inadequate or Imprecise Result or Readings	08/08/2023	08/08/2023
					CT2407	Active implantable	E2403	No Clinical Signs, Symptoms or Conditions	A0908	Incorrect, Inadequate or Imprecise Result or Readings	08/08/2023	08/08/2023
							F08	Hospitalization or Prolonged Hospitalization	A0908	Incorrect, Inadequate or Imprecise Result or Readings	08/08/2023	08/08/2023
							F1905	Device Revision or Replacement	A0908	Incorrect, Inadequate or Imprecise Result or Readings	08/08/2023	08/08/2023
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A0908	Incorrect, Inadequate or Imprecise Result or Readings	08/08/2023	08/08/2023
							F08	Hospitalization or Prolonged Hospitalization	A0908	Incorrect, Inadequate or Imprecise Result or Readings	08/08/2023	08/08/2023
							F1905	Device Revision or Replacement	A0908	Incorrect, Inadequate or Imprecise Result or Readings	08/08/2023	08/08/2023
					CT2727	Active implant devices/ancillaries	E2403	No Clinical Signs, Symptoms or Conditions	A0908	Incorrect, Inadequate or Imprecise Result or Readings	08/08/2023	08/08/2023
							F08	Hospitalization or Prolonged Hospitalization	A0908	Incorrect, Inadequate or Imprecise Result or Readings	08/08/2023	08/08/2023
							F1905	Device Revision or Replacement	A0908	Incorrect, Inadequate or Imprecise Result or Readings	08/08/2023	08/08/2023
					CT2750	Implantable cardiac electrical stimulators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A0908	Incorrect, Inadequate or Imprecise Result or Readings	08/08/2023	08/08/2023
							F08	Hospitalization or Prolonged Hospitalization	A0908	Incorrect, Inadequate or Imprecise Result or Readings	08/08/2023	08/08/2023
							F1905	Device Revision or Replacement	A0908	Incorrect, Inadequate or Imprecise Result or Readings	08/08/2023	08/08/2023
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A0908	Incorrect, Inadequate or Imprecise Result or Readings	08/08/2023	08/08/2023
							F08	Hospitalization or Prolonged Hospitalization	A0908	Incorrect, Inadequate or Imprecise Result or Readings	08/08/2023	08/08/2023
							F1905	Device Revision or Replacement	A0908	Incorrect, Inadequate or Imprecise Result or Readings	08/08/2023	08/08/2023
					CT2985	Active	E2403	No Clinical Signs, Symptoms or Conditions	A0908	Incorrect, Inadequate or Imprecise Result or Readings	08/08/2023	08/08/2023
							F08	Hospitalization or Prolonged Hospitalization	A0908	Incorrect, Inadequate or Imprecise Result or Readings	08/08/2023	08/08/2023
							F1905	Device Revision or Replacement	A0908	Incorrect, Inadequate or Imprecise Result or Readings	08/08/2023	08/08/2023
					CT3001	Implantable metal	E2403	No Clinical Signs, Symptoms or Conditions	A0908	Incorrect, Inadequate or Imprecise Result or Readings	08/08/2023	08/08/2023
							F08	Hospitalization or Prolonged Hospitalization	A0908	Incorrect, Inadequate or Imprecise Result or Readings	08/08/2023	08/08/2023
							F1905	Device Revision or Replacement	A0908	Incorrect, Inadequate or Imprecise Result or Readings	08/08/2023	08/08/2023

					CT3002	Implantable synthetic polymer	E2403	No Clinical Signs, Symptoms or Conditions	A0908	Incorrect, Inadequate or Imprecise Result or Readings	08/08/2023	08/08/2023
							F08	Hospitalization or Prolonged Hospitalization	A0908	Incorrect, Inadequate or Imprecise Result or Readings	08/08/2023	08/08/2023
							F1905	Device Revision or Replacement	A0908	Incorrect, Inadequate or Imprecise Result or Readings	08/08/2023	08/08/2023
					CT3006	Implantable inorganic material	E2403	No Clinical Signs, Symptoms or Conditions	A0908	Incorrect, Inadequate or Imprecise Result or Readings	08/08/2023	08/08/2023
							F08	Hospitalization or Prolonged Hospitalization	A0908	Incorrect, Inadequate or Imprecise Result or Readings	08/08/2023	08/08/2023
							F1905	Device Revision or Replacement	A0908	Incorrect, Inadequate or Imprecise Result or Readings	08/08/2023	08/08/2023
					CT314	Electrically powered	E2403	No Clinical Signs, Symptoms or Conditions	A0908	Incorrect, Inadequate or Imprecise Result or Readings	08/08/2023	08/08/2023
							F08	Hospitalization or Prolonged Hospitalization	A0908	Incorrect, Inadequate or Imprecise Result or Readings	08/08/2023	08/08/2023
							F1905	Device Revision or Replacement	A0908	Incorrect, Inadequate or Imprecise Result or Readings	08/08/2023	08/08/2023
					CT321	Long-term surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A0908	Incorrect, Inadequate or Imprecise Result or Readings	08/08/2023	08/08/2023
							F08	Hospitalization or Prolonged Hospitalization	A0908	Incorrect, Inadequate or Imprecise Result or Readings	08/08/2023	08/08/2023
							F1905	Device Revision or Replacement	A0908	Incorrect, Inadequate or Imprecise Result or Readings	08/08/2023	08/08/2023
					CT449	Generators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A0908	Incorrect, Inadequate or Imprecise Result or Readings	08/08/2023	08/08/2023
							F08	Hospitalization or Prolonged Hospitalization	A0908	Incorrect, Inadequate or Imprecise Result or Readings	08/08/2023	08/08/2023
							F1905	Device Revision or Replacement	A0908	Incorrect, Inadequate or Imprecise Result or Readings	08/08/2023	08/08/2023
					CT981	Single use	E2403	No Clinical Signs, Symptoms or Conditions	A0908	Incorrect, Inadequate or Imprecise Result or Readings	08/08/2023	08/08/2023
							F08	Hospitalization or Prolonged Hospitalization	A0908	Incorrect, Inadequate or Imprecise Result or Readings	08/08/2023	08/08/2023
							F1905	Device Revision or Replacement	A0908	Incorrect, Inadequate or Imprecise Result or Readings	08/08/2023	08/08/2023
					CT983	Surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A0908	Incorrect, Inadequate or Imprecise Result or Readings	08/08/2023	08/08/2023
							F08	Hospitalization or Prolonged Hospitalization	A0908	Incorrect, Inadequate or Imprecise Result or Readings	08/08/2023	08/08/2023
							F1905	Device Revision or Replacement	A0908	Incorrect, Inadequate or Imprecise Result or Readings	08/08/2023	08/08/2023
14528	2023	All other reportable incidents	35223	Endocardial/interventricular septal pacing lead	CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A07	Electrical /Electronic Property Problem		
									A072202	Low Impedance		
							F26	No Health Consequences or Impact	A07	Electrical /Electronic Property Problem		
									A072202	Low Impedance		
14529	2023		35822	External pacemaker, transcutaneous pacing	CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A08	Calibration Problem		
									A23	Use of Device Problem		
							F23	Unexpected Medical Intervention	A08	Calibration Problem		
									A23	Use of Device Problem		
							F26	No Health Consequences or Impact	A08	Calibration Problem		
									A23	Use of Device Problem		
14530	2023	All other reportable incidents	35852	Single-chamber implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E2104	Electric Shock	A0709	Device Sensing Problem		
									A070909	Oversensing		
									A071302	Inappropriate/Inadequate Shock/Stimulation		
									A1301	Failure to Read Input Signal		
							F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A0709	Device Sensing Problem		
									A070909	Oversensing		
									A071302	Inappropriate/Inadequate Shock/Stimulation		
									A1301	Failure to Read Input Signal		
							F2203	Imaging Required	A0709	Device Sensing Problem		
									A070909	Oversensing		
									A071302	Inappropriate/Inadequate Shock/Stimulation		
									A1301	Failure to Read Input Signal		
14531	2023	All other reportable incidents	47265	Dual-chamber implantable pacemaker, rate-responsive	CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A070909	Oversensing		
									A0712	Pacing Problem		
									A090801	Signal Artifact/Noise		
							F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A070909	Oversensing		
									A0712	Pacing Problem		
									A090801	Signal Artifact/Noise		
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A070909	Oversensing		
									A0712	Pacing Problem		
									A090801	Signal Artifact/Noise		
							F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A070909	Oversensing		
									A0712	Pacing Problem		
									A090801	Signal Artifact/Noise		
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A070909	Oversensing		
									A0712	Pacing Problem		
									A090801	Signal Artifact/Noise		
							F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A070909	Oversensing		
									A0712	Pacing Problem		
									A090801	Signal Artifact/Noise		
14532	2023	All other reportable incidents	47205	Cardiac pulse generator programmer	CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A150105	Key or Button Unresponsive/not Working		
							F26	No Health Consequences or Impact	A150105	Key or Button Unresponsive/not Working		
14533	2023	All other reportable incidents	47265	Dual-chamber implantable pacemaker, rate-responsive	CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A0908	Incorrect, Inadequate or Imprecise Result or Readings	17/11/2023	17/11/2023
							F08	Hospitalization or Prolonged Hospitalization	A0908	Incorrect, Inadequate or Imprecise Result or Readings	17/11/2023	17/11/2023
							F1905	Device Revision or Replacement	A0908	Incorrect, Inadequate or Imprecise Result or Readings	17/11/2023	17/11/2023
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A0908	Incorrect, Inadequate or Imprecise Result or Readings	17/11/2023	17/11/2023
							F08	Hospitalization or Prolonged Hospitalization	A0908	Incorrect, Inadequate or Imprecise Result or Readings	17/11/2023	17/11/2023

						F1905	Device Revision or Replacement	A0908	Incorrect, Inadequate or Imprecise Result or Readings	17/11/2023	17/11/2023
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A0908	Incorrect, Inadequate or Imprecise Result or Readings	17/11/2023	17/11/2023
							F08	Hospitalization or Prolonged Hospitalization	A0908	Incorrect, Inadequate or Imprecise Result or Readings	17/11/2023	17/11/2023
							F1905	Device Revision or Replacement	A0908	Incorrect, Inadequate or Imprecise Result or Readings	17/11/2023	17/11/2023
14534	2023	All other reportable incidents	37265	Dual-chamber implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E2104	Electric Shock	A070102	High Capture Threshold		
									A070909	Oversensing		
									A0712	Pacing Problem		
									A071302	Inappropriate/Inadequate Shock/Stimulation		
									A0722	Impedance Problem		
									A1301	Failure to Read Input Signal		
							F08	Hospitalization or Prolonged Hospitalization	A070102	High Capture Threshold		
									A070909	Oversensing		
									A0712	Pacing Problem		
									A071302	Inappropriate/Inadequate Shock/Stimulation		
									A0722	Impedance Problem		
									A1301	Failure to Read Input Signal		
							F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A070102	High Capture Threshold		
									A070909	Oversensing		
									A0712	Pacing Problem		
									A071302	Inappropriate/Inadequate Shock/Stimulation		
									A0722	Impedance Problem		
									A1301	Failure to Read Input Signal		
14535	2023	All other reportable incidents	35852	Single-chamber implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
14536	2023	All other reportable incidents	47265	Dual-chamber implantable pacemaker, rate-responsive	CT141	Electrodes/Leads and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A0908	Incorrect, Inadequate or Imprecise Result or Readings	21/11/2023	21/11/2023
							F08	Hospitalization or Prolonged Hospitalization	A0908	Incorrect, Inadequate or Imprecise Result or Readings	21/11/2023	21/11/2023
							F1905	Device Revision or Replacement	A0908	Incorrect, Inadequate or Imprecise Result or Readings	21/11/2023	21/11/2023
					CT145	Cardiovascular devices	E2403	No Clinical Signs, Symptoms or Conditions	A0908	Incorrect, Inadequate or Imprecise Result or Readings	21/11/2023	21/11/2023
							F08	Hospitalization or Prolonged Hospitalization	A0908	Incorrect, Inadequate or Imprecise Result or Readings	21/11/2023	21/11/2023
							F1905	Device Revision or Replacement	A0908	Incorrect, Inadequate or Imprecise Result or Readings	21/11/2023	21/11/2023
					CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A0908	Incorrect, Inadequate or Imprecise Result or Readings	21/11/2023	21/11/2023
							F08	Hospitalization or Prolonged Hospitalization	A0908	Incorrect, Inadequate or Imprecise Result or Readings	21/11/2023	21/11/2023
							F1905	Device Revision or Replacement	A0908	Incorrect, Inadequate or Imprecise Result or Readings	21/11/2023	21/11/2023
					CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A0908	Incorrect, Inadequate or Imprecise Result or Readings	21/11/2023	21/11/2023
							F08	Hospitalization or Prolonged Hospitalization	A0908	Incorrect, Inadequate or Imprecise Result or Readings	21/11/2023	21/11/2023
							F1905	Device Revision or Replacement	A0908	Incorrect, Inadequate or Imprecise Result or Readings	21/11/2023	21/11/2023
					CT233	Surgical	E2403	No Clinical Signs, Symptoms or Conditions	A0908	Incorrect, Inadequate or Imprecise Result or Readings	21/11/2023	21/11/2023
							F08	Hospitalization or Prolonged Hospitalization	A0908	Incorrect, Inadequate or Imprecise Result or Readings	21/11/2023	21/11/2023
							F1905	Device Revision or Replacement	A0908	Incorrect, Inadequate or Imprecise Result or Readings	21/11/2023	21/11/2023
					CT2406	Implantable	E2403	No Clinical Signs, Symptoms or Conditions	A0908	Incorrect, Inadequate or Imprecise Result or Readings	21/11/2023	21/11/2023
							F08	Hospitalization or Prolonged Hospitalization	A0908	Incorrect, Inadequate or Imprecise Result or Readings	21/11/2023	21/11/2023
							F1905	Device Revision or Replacement	A0908	Incorrect, Inadequate or Imprecise Result or Readings	21/11/2023	21/11/2023
					CT2407	Active implantable	E2403	No Clinical Signs, Symptoms or Conditions	A0908	Incorrect, Inadequate or Imprecise Result or Readings	21/11/2023	21/11/2023
							F08	Hospitalization or Prolonged Hospitalization	A0908	Incorrect, Inadequate or Imprecise Result or Readings	21/11/2023	21/11/2023
							F1905	Device Revision or Replacement	A0908	Incorrect, Inadequate or Imprecise Result or Readings	21/11/2023	21/11/2023
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A0908	Incorrect, Inadequate or Imprecise Result or Readings	21/11/2023	21/11/2023
							F08	Hospitalization or Prolonged Hospitalization	A0908	Incorrect, Inadequate or Imprecise Result or Readings	21/11/2023	21/11/2023
							F1905	Device Revision or Replacement	A0908	Incorrect, Inadequate or Imprecise Result or Readings	21/11/2023	21/11/2023
					CT2727	Active implant devices/ancillaries	E2403	No Clinical Signs, Symptoms or Conditions	A0908	Incorrect, Inadequate or Imprecise Result or Readings	21/11/2023	21/11/2023
							F08	Hospitalization or Prolonged Hospitalization	A0908	Incorrect, Inadequate or Imprecise Result or Readings	21/11/2023	21/11/2023
							F1905	Device Revision or Replacement	A0908	Incorrect, Inadequate or Imprecise Result or Readings	21/11/2023	21/11/2023
					CT2750	Implantable cardiac electrical stimulators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A0908	Incorrect, Inadequate or Imprecise Result or Readings	21/11/2023	21/11/2023
							F08	Hospitalization or Prolonged Hospitalization	A0908	Incorrect, Inadequate or Imprecise Result or Readings	21/11/2023	21/11/2023
							F1905	Device Revision or Replacement	A0908	Incorrect, Inadequate or Imprecise Result or Readings	21/11/2023	21/11/2023
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A0908	Incorrect, Inadequate or Imprecise Result or Readings	21/11/2023	21/11/2023
							F08	Hospitalization or Prolonged Hospitalization	A0908	Incorrect, Inadequate or Imprecise Result or Readings	21/11/2023	21/11/2023
							F1905	Device Revision or Replacement	A0908	Incorrect, Inadequate or Imprecise Result or Readings	21/11/2023	21/11/2023

					CT2985	Active	E2403	No Clinical Signs, Symptoms or Conditions	A0908	Incorrect, Inadequate or Imprecise Result or Readings	21/11/2023	21/11/2023
							F08	Hospitalization or Prolonged Hospitalization	A0908	Incorrect, Inadequate or Imprecise Result or Readings	21/11/2023	21/11/2023
							F1905	Device Revision or Replacement	A0908	Incorrect, Inadequate or Imprecise Result or Readings	21/11/2023	21/11/2023
					CT3001	Implantable metal	E2403	No Clinical Signs, Symptoms or Conditions	A0908	Incorrect, Inadequate or Imprecise Result or Readings	21/11/2023	21/11/2023
							F08	Hospitalization or Prolonged Hospitalization	A0908	Incorrect, Inadequate or Imprecise Result or Readings	21/11/2023	21/11/2023
							F1905	Device Revision or Replacement	A0908	Incorrect, Inadequate or Imprecise Result or Readings	21/11/2023	21/11/2023
					CT3002	Implantable synthetic polymer	E2403	No Clinical Signs, Symptoms or Conditions	A0908	Incorrect, Inadequate or Imprecise Result or Readings	21/11/2023	21/11/2023
							F08	Hospitalization or Prolonged Hospitalization	A0908	Incorrect, Inadequate or Imprecise Result or Readings	21/11/2023	21/11/2023
							F1905	Device Revision or Replacement	A0908	Incorrect, Inadequate or Imprecise Result or Readings	21/11/2023	21/11/2023
					CT3006	Implantable inorganic material	E2403	No Clinical Signs, Symptoms or Conditions	A0908	Incorrect, Inadequate or Imprecise Result or Readings	21/11/2023	21/11/2023
							F08	Hospitalization or Prolonged Hospitalization	A0908	Incorrect, Inadequate or Imprecise Result or Readings	21/11/2023	21/11/2023
							F1905	Device Revision or Replacement	A0908	Incorrect, Inadequate or Imprecise Result or Readings	21/11/2023	21/11/2023
					CT314	Electrically powered	E2403	No Clinical Signs, Symptoms or Conditions	A0908	Incorrect, Inadequate or Imprecise Result or Readings	21/11/2023	21/11/2023
							F08	Hospitalization or Prolonged Hospitalization	A0908	Incorrect, Inadequate or Imprecise Result or Readings	21/11/2023	21/11/2023
							F1905	Device Revision or Replacement	A0908	Incorrect, Inadequate or Imprecise Result or Readings	21/11/2023	21/11/2023
					CT321	Long-term surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A0908	Incorrect, Inadequate or Imprecise Result or Readings	21/11/2023	21/11/2023
							F08	Hospitalization or Prolonged Hospitalization	A0908	Incorrect, Inadequate or Imprecise Result or Readings	21/11/2023	21/11/2023
							F1905	Device Revision or Replacement	A0908	Incorrect, Inadequate or Imprecise Result or Readings	21/11/2023	21/11/2023
					CT449	Generators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A0908	Incorrect, Inadequate or Imprecise Result or Readings	21/11/2023	21/11/2023
							F08	Hospitalization or Prolonged Hospitalization	A0908	Incorrect, Inadequate or Imprecise Result or Readings	21/11/2023	21/11/2023
							F1905	Device Revision or Replacement	A0908	Incorrect, Inadequate or Imprecise Result or Readings	21/11/2023	21/11/2023
					CT981	Single use	E2403	No Clinical Signs, Symptoms or Conditions	A0908	Incorrect, Inadequate or Imprecise Result or Readings	21/11/2023	21/11/2023
							F08	Hospitalization or Prolonged Hospitalization	A0908	Incorrect, Inadequate or Imprecise Result or Readings	21/11/2023	21/11/2023
							F1905	Device Revision or Replacement	A0908	Incorrect, Inadequate or Imprecise Result or Readings	21/11/2023	21/11/2023
					CT983	Surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A0908	Incorrect, Inadequate or Imprecise Result or Readings	21/11/2023	21/11/2023
							F08	Hospitalization or Prolonged Hospitalization	A0908	Incorrect, Inadequate or Imprecise Result or Readings	21/11/2023	21/11/2023
							F1905	Device Revision or Replacement	A0908	Incorrect, Inadequate or Imprecise Result or Readings	21/11/2023	21/11/2023
14537	2023	All other reportable incidents	47263	Cardiac resynchronization therapy implantable pacemaker	CT1685	Cardiac pacemakers	E2103	Device Overstimulation of Tissue	A0908	Incorrect, Inadequate or Imprecise Result or Readings	23/11/2023	23/11/2023
							F08	Hospitalization or Prolonged Hospitalization	A0908	Incorrect, Inadequate or Imprecise Result or Readings	23/11/2023	23/11/2023
							F1905	Device Revision or Replacement	A0908	Incorrect, Inadequate or Imprecise Result or Readings	23/11/2023	23/11/2023
					CT257	Implantable cardiac pacemakers	E2103	Device Overstimulation of Tissue	A0908	Incorrect, Inadequate or Imprecise Result or Readings	23/11/2023	23/11/2023
							F08	Hospitalization or Prolonged Hospitalization	A0908	Incorrect, Inadequate or Imprecise Result or Readings	23/11/2023	23/11/2023
							F1905	Device Revision or Replacement	A0908	Incorrect, Inadequate or Imprecise Result or Readings	23/11/2023	23/11/2023
					CT2866	Implantable cardiac electrical stimulators	E2103	Device Overstimulation of Tissue	A0908	Incorrect, Inadequate or Imprecise Result or Readings	23/11/2023	23/11/2023
							F08	Hospitalization or Prolonged Hospitalization	A0908	Incorrect, Inadequate or Imprecise Result or Readings	23/11/2023	23/11/2023
							F1905	Device Revision or Replacement	A0908	Incorrect, Inadequate or Imprecise Result or Readings	23/11/2023	23/11/2023
14538	2023		47265	Dual-chamber implantable pacemaker, rate-responsive	CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	27/11/2023	27/11/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	27/11/2023	27/11/2023
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	27/11/2023	27/11/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	27/11/2023	27/11/2023
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	27/11/2023	27/11/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	27/11/2023	27/11/2023
14539	2023		47265	Dual-chamber implantable pacemaker, rate-responsive	CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	23/11/2023	23/11/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	23/11/2023	23/11/2023
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	23/11/2023	23/11/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	23/11/2023	23/11/2023
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	23/11/2023	23/11/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	23/11/2023	23/11/2023
14540	2023	All other reportable incidents	17882	Physiologic-monitoring defibrillation system	CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070802	Intermittent Loss of Power		
							F27	Problem Identified During Non-Clinical Procedure	A070802	Intermittent Loss of Power		
14541	2023	All other reportable incidents	47265	Dual-chamber implantable pacemaker, rate-responsive	CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	28/11/2023	28/11/2023
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	28/11/2023	28/11/2023
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	28/11/2023	28/11/2023
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	28/11/2023	28/11/2023
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	28/11/2023	28/11/2023
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	28/11/2023	28/11/2023
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	28/11/2023	28/11/2023
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	28/11/2023	28/11/2023
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	28/11/2023	28/11/2023
14542	2023	All other reportable incidents	47263	Cardiac resynchronization therapy implantable pacemaker	CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A0908	Incorrect, Inadequate or Imprecise Result or Readings	17/03/2023	17/03/2023

							F08	Hospitalization or Prolonged Hospitalization	A0908	Incorrect, Inadequate or Imprecise Result or Readings	17/03/2023	17/03/2023
							F1905	Device Revision or Replacement	A0908	Incorrect, Inadequate or Imprecise Result or Readings	17/03/2023	17/03/2023
					CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A0908	Incorrect, Inadequate or Imprecise Result or Readings	17/03/2023	17/03/2023
							F08	Hospitalization or Prolonged Hospitalization	A0908	Incorrect, Inadequate or Imprecise Result or Readings	17/03/2023	17/03/2023
							F1905	Device Revision or Replacement	A0908	Incorrect, Inadequate or Imprecise Result or Readings	17/03/2023	17/03/2023
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A0908	Incorrect, Inadequate or Imprecise Result or Readings	17/03/2023	17/03/2023
							F08	Hospitalization or Prolonged Hospitalization	A0908	Incorrect, Inadequate or Imprecise Result or Readings	17/03/2023	17/03/2023
							F1905	Device Revision or Replacement	A0908	Incorrect, Inadequate or Imprecise Result or Readings	17/03/2023	17/03/2023
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A0908	Incorrect, Inadequate or Imprecise Result or Readings	17/03/2023	17/03/2023
							F08	Hospitalization or Prolonged Hospitalization	A0908	Incorrect, Inadequate or Imprecise Result or Readings	17/03/2023	17/03/2023
							F1905	Device Revision or Replacement	A0908	Incorrect, Inadequate or Imprecise Result or Readings	17/03/2023	17/03/2023
14543	2023	All other reportable incidents	35852	Single-chamber implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	09/01/2024	09/01/2024
									A0908	Incorrect, Inadequate or Imprecise Result or Readings	09/01/2024	09/01/2024
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	09/01/2024	09/01/2024
									A0908	Incorrect, Inadequate or Imprecise Result or Readings	09/01/2024	09/01/2024
							F1903	Device Explantation	A070504	Premature Discharge of Battery	09/01/2024	09/01/2024
									A0908	Incorrect, Inadequate or Imprecise Result or Readings	09/01/2024	09/01/2024
14544	2023	All other reportable incidents	35852	Single-chamber implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
14545	2023	All other reportable incidents	37265	Dual-chamber implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	23/06/2023	23/06/2023
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	23/06/2023	23/06/2023
							F12	Serious Injury/ Illness/ Impairment	A070504	Premature Discharge of Battery	23/06/2023	23/06/2023
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	23/06/2023	23/06/2023
14546	2023	Death	47265	Dual-chamber implantable pacemaker, rate-responsive	CT141	Electrodes/Leads and associated devices	E2401	Insufficient Information	A070101	Failure to Capture	28/11/2023	28/11/2023
									A070102	High Capture Threshold	28/11/2023	28/11/2023
									A070504	Premature Discharge of Battery	28/11/2023	28/11/2023
							F02	Death	A070101	Failure to Capture	28/11/2023	28/11/2023
									A070102	High Capture Threshold	28/11/2023	28/11/2023
									A070504	Premature Discharge of Battery	28/11/2023	28/11/2023
							F08	Hospitalization or Prolonged Hospitalization	A070101	Failure to Capture	28/11/2023	28/11/2023
									A070102	High Capture Threshold	28/11/2023	28/11/2023
									A070504	Premature Discharge of Battery	28/11/2023	28/11/2023
					CT233	Surgical	E2401	Insufficient Information	A070101	Failure to Capture	28/11/2023	28/11/2023
									A070102	High Capture Threshold	28/11/2023	28/11/2023
									A070504	Premature Discharge of Battery	28/11/2023	28/11/2023
							F02	Death	A070101	Failure to Capture	28/11/2023	28/11/2023
									A070102	High Capture Threshold	28/11/2023	28/11/2023
									A070504	Premature Discharge of Battery	28/11/2023	28/11/2023
							F08	Hospitalization or Prolonged Hospitalization	A070101	Failure to Capture	28/11/2023	28/11/2023
									A070102	High Capture Threshold	28/11/2023	28/11/2023
									A070504	Premature Discharge of Battery	28/11/2023	28/11/2023
					CT2407	Active implantable	E2401	Insufficient Information	A070101	Failure to Capture	28/11/2023	28/11/2023
									A070102	High Capture Threshold	28/11/2023	28/11/2023
									A070504	Premature Discharge of Battery	28/11/2023	28/11/2023
							F02	Death	A070101	Failure to Capture	28/11/2023	28/11/2023
									A070102	High Capture Threshold	28/11/2023	28/11/2023
									A070504	Premature Discharge of Battery	28/11/2023	28/11/2023
							F08	Hospitalization or Prolonged Hospitalization	A070101	Failure to Capture	28/11/2023	28/11/2023
									A070102	High Capture Threshold	28/11/2023	28/11/2023
									A070504	Premature Discharge of Battery	28/11/2023	28/11/2023
					CT257	Implantable cardiac pacemakers	E2401	Insufficient Information	A070101	Failure to Capture	28/11/2023	28/11/2023
									A070102	High Capture Threshold	28/11/2023	28/11/2023
									A070504	Premature Discharge of Battery	28/11/2023	28/11/2023
							F02	Death	A070101	Failure to Capture	28/11/2023	28/11/2023
									A070102	High Capture Threshold	28/11/2023	28/11/2023
									A070504	Premature Discharge of Battery	28/11/2023	28/11/2023
							F08	Hospitalization or Prolonged Hospitalization	A070101	Failure to Capture	28/11/2023	28/11/2023
									A070102	High Capture Threshold	28/11/2023	28/11/2023
									A070504	Premature Discharge of Battery	28/11/2023	28/11/2023
					CT3001	Implantable metal	E2401	Insufficient Information	A070101	Failure to Capture	28/11/2023	28/11/2023
									A070102	High Capture Threshold	28/11/2023	28/11/2023
									A070504	Premature Discharge of Battery	28/11/2023	28/11/2023
							F02	Death	A070101	Failure to Capture	28/11/2023	28/11/2023
									A070102	High Capture Threshold	28/11/2023	28/11/2023
									A070504	Premature Discharge of Battery	28/11/2023	28/11/2023
							F08	Hospitalization or Prolonged Hospitalization	A070101	Failure to Capture	28/11/2023	28/11/2023
									A070102	High Capture Threshold	28/11/2023	28/11/2023
									A070504	Premature Discharge of Battery	28/11/2023	28/11/2023
					CT3002	Implantable synthetic polymer	E2401	Insufficient Information	A070101	Failure to Capture	28/11/2023	28/11/2023
									A070102	High Capture Threshold	28/11/2023	28/11/2023
									A070504	Premature Discharge of Battery	28/11/2023	28/11/2023
							F02	Death	A070101	Failure to Capture	28/11/2023	28/11/2023
									A070102	High Capture Threshold	28/11/2023	28/11/2023
									A070504	Premature Discharge of Battery	28/11/2023	28/11/2023
							F08	Hospitalization or Prolonged Hospitalization	A070101	Failure to Capture	28/11/2023	28/11/2023
									A070102	High Capture Threshold	28/11/2023	28/11/2023
									A070504	Premature Discharge of Battery	28/11/2023	28/11/2023

					CT314	Electrically powered	E2401	Insufficient Information	A070101	Failure to Capture	28/11/2023	28/11/2023
									A070102	High Capture Threshold	28/11/2023	28/11/2023
									A070504	Premature Discharge of Battery	28/11/2023	28/11/2023
							F02	Death	A070101	Failure to Capture	28/11/2023	28/11/2023
									A070102	High Capture Threshold	28/11/2023	28/11/2023
									A070504	Premature Discharge of Battery	28/11/2023	28/11/2023
							F08	Hospitalization or Prolonged Hospitalization	A070101	Failure to Capture	28/11/2023	28/11/2023
									A070102	High Capture Threshold	28/11/2023	28/11/2023
									A070504	Premature Discharge of Battery	28/11/2023	28/11/2023
					CT321	Long-term surgical invasive	E2401	Insufficient Information	A070101	Failure to Capture	28/11/2023	28/11/2023
									A070102	High Capture Threshold	28/11/2023	28/11/2023
									A070504	Premature Discharge of Battery	28/11/2023	28/11/2023
							F02	Death	A070101	Failure to Capture	28/11/2023	28/11/2023
									A070102	High Capture Threshold	28/11/2023	28/11/2023
									A070504	Premature Discharge of Battery	28/11/2023	28/11/2023
							F08	Hospitalization or Prolonged Hospitalization	A070101	Failure to Capture	28/11/2023	28/11/2023
									A070102	High Capture Threshold	28/11/2023	28/11/2023
									A070504	Premature Discharge of Battery	28/11/2023	28/11/2023
					CT449	Generators and associated devices	E2401	Insufficient Information	A070101	Failure to Capture	28/11/2023	28/11/2023
									A070102	High Capture Threshold	28/11/2023	28/11/2023
									A070504	Premature Discharge of Battery	28/11/2023	28/11/2023
							F02	Death	A070101	Failure to Capture	28/11/2023	28/11/2023
									A070102	High Capture Threshold	28/11/2023	28/11/2023
									A070504	Premature Discharge of Battery	28/11/2023	28/11/2023
							F08	Hospitalization or Prolonged Hospitalization	A070101	Failure to Capture	28/11/2023	28/11/2023
									A070102	High Capture Threshold	28/11/2023	28/11/2023
									A070504	Premature Discharge of Battery	28/11/2023	28/11/2023
					CT981	Single use	E2401	Insufficient Information	A070101	Failure to Capture	28/11/2023	28/11/2023
									A070102	High Capture Threshold	28/11/2023	28/11/2023
									A070504	Premature Discharge of Battery	28/11/2023	28/11/2023
							F02	Death	A070101	Failure to Capture	28/11/2023	28/11/2023
									A070102	High Capture Threshold	28/11/2023	28/11/2023
									A070504	Premature Discharge of Battery	28/11/2023	28/11/2023
							F08	Hospitalization or Prolonged Hospitalization	A070101	Failure to Capture	28/11/2023	28/11/2023
									A070102	High Capture Threshold	28/11/2023	28/11/2023
									A070504	Premature Discharge of Battery	28/11/2023	28/11/2023
14547	2023		47265	Dual-chamber implantable pacemaker, rate-responsive	CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A0203	Defective Device	28/11/2023	28/11/2023
									A070909	Oversensing	28/11/2023	28/11/2023
							F1901	Additional Surgery	A0203	Defective Device	28/11/2023	28/11/2023
									A070909	Oversensing	28/11/2023	28/11/2023
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A0203	Defective Device	28/11/2023	28/11/2023
									A070909	Oversensing	28/11/2023	28/11/2023
							F1901	Additional Surgery	A0203	Defective Device	28/11/2023	28/11/2023
									A070909	Oversensing	28/11/2023	28/11/2023
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A0203	Defective Device	28/11/2023	28/11/2023
									A070909	Oversensing	28/11/2023	28/11/2023
							F1901	Additional Surgery	A0203	Defective Device	28/11/2023	28/11/2023
									A070909	Oversensing	28/11/2023	28/11/2023
14548	2023	All other reportable incidents	35852	Single-chamber implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E2104	Electric Shock	A070909	Oversensing		
									A071302	Inappropriate/Inadequate Shock/Stimulation		
									A090801	Signal Artifact/Noise		
							F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A070909	Oversensing		
									A071302	Inappropriate/Inadequate Shock/Stimulation		
									A090801	Signal Artifact/Noise		
							F2203	Imaging Required	A070909	Oversensing		
									A071302	Inappropriate/Inadequate Shock/Stimulation		
									A090801	Signal Artifact/Noise		
14549	2023	All other reportable incidents	35852	Single-chamber implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E2104	Electric Shock	A070909	Oversensing		
									A071302	Inappropriate/Inadequate Shock/Stimulation		
							F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A070909	Oversensing		
									A071302	Inappropriate/Inadequate Shock/Stimulation		
							F15	Recognized Device or Procedural Complication	A070909	Oversensing		
									A071302	Inappropriate/Inadequate Shock/Stimulation		
14550	2023	All other reportable incidents	47265	Dual-chamber implantable pacemaker, rate-responsive	CT1685	Cardiac pacemakers	E0112	Dizziness	A071102	Failure to Interrogate	27/11/2023	27/11/2023
									A071204	No Pacing	27/11/2023	27/11/2023
							F1901	Additional Surgery	A071102	Failure to Interrogate	27/11/2023	27/11/2023
									A071204	No Pacing	27/11/2023	27/11/2023
					CT257	Implantable cardiac pacemakers	E0112	Dizziness	A071102	Failure to Interrogate	27/11/2023	27/11/2023
									A071204	No Pacing	27/11/2023	27/11/2023
							F1901	Additional Surgery	A071102	Failure to Interrogate	27/11/2023	27/11/2023
									A071204	No Pacing	27/11/2023	27/11/2023
					CT2866	Implantable cardiac electrical stimulators	E0112	Dizziness	A071102	Failure to Interrogate	27/11/2023	27/11/2023
									A071204	No Pacing	27/11/2023	27/11/2023
							F1901	Additional Surgery	A071102	Failure to Interrogate	27/11/2023	27/11/2023
									A071204	No Pacing	27/11/2023	27/11/2023
14551	2023	All other reportable incidents	47267	Single-chamber implantable pacemaker, rate-responsive	CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A070909	Oversensing		
									A072201	High Impedance		
							F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A070909	Oversensing		
									A072201	High Impedance		
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A070909	Oversensing		
									A072201	High Impedance		
							F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A070909	Oversensing		
									A072201	High Impedance		
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A070909	Oversensing		
									A072201	High Impedance		
							F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A070909	Oversensing		
									A072201	High Impedance		
14552	2023	All other reportable incidents	47265	Dual-chamber implantable pacemaker, rate-	CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	07/09/2022	07/09/2022

				responsive			F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	07/09/2022	07/09/2022
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	07/09/2022	07/09/2022
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	07/09/2022	07/09/2022
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	07/09/2022	07/09/2022
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	07/09/2022	07/09/2022
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	07/09/2022	07/09/2022
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	07/09/2022	07/09/2022
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	07/09/2022	07/09/2022
14553	2023	All other reportable incidents	47267	Single-chamber implantable pacemaker, rate-responsive	CT141	Electrodes/Leads and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A0203	Defective Device	21/01/2023	21/01/2023
									A070504	Premature Discharge of Battery	21/01/2023	21/01/2023
							F08	Hospitalization or Prolonged Hospitalization	A0203	Defective Device	21/01/2023	21/01/2023
									A070504	Premature Discharge of Battery	21/01/2023	21/01/2023
							F1905	Device Revision or Replacement	A0203	Defective Device	21/01/2023	21/01/2023
									A070504	Premature Discharge of Battery	21/01/2023	21/01/2023
					CT145	Cardiovascular devices	E2403	No Clinical Signs, Symptoms or Conditions	A0203	Defective Device	21/01/2023	21/01/2023
									A070504	Premature Discharge of Battery	21/01/2023	21/01/2023
							F08	Hospitalization or Prolonged Hospitalization	A0203	Defective Device	21/01/2023	21/01/2023
									A070504	Premature Discharge of Battery	21/01/2023	21/01/2023
							F1905	Device Revision or Replacement	A0203	Defective Device	21/01/2023	21/01/2023
									A070504	Premature Discharge of Battery	21/01/2023	21/01/2023
					CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A0203	Defective Device	21/01/2023	21/01/2023
									A070504	Premature Discharge of Battery	21/01/2023	21/01/2023
							F08	Hospitalization or Prolonged Hospitalization	A0203	Defective Device	21/01/2023	21/01/2023
									A070504	Premature Discharge of Battery	21/01/2023	21/01/2023
							F1905	Device Revision or Replacement	A0203	Defective Device	21/01/2023	21/01/2023
									A070504	Premature Discharge of Battery	21/01/2023	21/01/2023
					CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A0203	Defective Device	21/01/2023	21/01/2023
									A070504	Premature Discharge of Battery	21/01/2023	21/01/2023
							F08	Hospitalization or Prolonged Hospitalization	A0203	Defective Device	21/01/2023	21/01/2023
									A070504	Premature Discharge of Battery	21/01/2023	21/01/2023
							F1905	Device Revision or Replacement	A0203	Defective Device	21/01/2023	21/01/2023
									A070504	Premature Discharge of Battery	21/01/2023	21/01/2023
					CT233	Surgical	E2403	No Clinical Signs, Symptoms or Conditions	A0203	Defective Device	21/01/2023	21/01/2023
									A070504	Premature Discharge of Battery	21/01/2023	21/01/2023
							F08	Hospitalization or Prolonged Hospitalization	A0203	Defective Device	21/01/2023	21/01/2023
									A070504	Premature Discharge of Battery	21/01/2023	21/01/2023
							F1905	Device Revision or Replacement	A0203	Defective Device	21/01/2023	21/01/2023
									A070504	Premature Discharge of Battery	21/01/2023	21/01/2023
					CT2406	Implantable	E2403	No Clinical Signs, Symptoms or Conditions	A0203	Defective Device	21/01/2023	21/01/2023
									A070504	Premature Discharge of Battery	21/01/2023	21/01/2023
							F08	Hospitalization or Prolonged Hospitalization	A0203	Defective Device	21/01/2023	21/01/2023
									A070504	Premature Discharge of Battery	21/01/2023	21/01/2023
							F1905	Device Revision or Replacement	A0203	Defective Device	21/01/2023	21/01/2023
									A070504	Premature Discharge of Battery	21/01/2023	21/01/2023
					CT2407	Active implantable	E2403	No Clinical Signs, Symptoms or Conditions	A0203	Defective Device	21/01/2023	21/01/2023
									A070504	Premature Discharge of Battery	21/01/2023	21/01/2023
							F08	Hospitalization or Prolonged Hospitalization	A0203	Defective Device	21/01/2023	21/01/2023
									A070504	Premature Discharge of Battery	21/01/2023	21/01/2023
							F1905	Device Revision or Replacement	A0203	Defective Device	21/01/2023	21/01/2023
									A070504	Premature Discharge of Battery	21/01/2023	21/01/2023
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A0203	Defective Device	21/01/2023	21/01/2023
									A070504	Premature Discharge of Battery	21/01/2023	21/01/2023
							F08	Hospitalization or Prolonged Hospitalization	A0203	Defective Device	21/01/2023	21/01/2023
									A070504	Premature Discharge of Battery	21/01/2023	21/01/2023
							F1905	Device Revision or Replacement	A0203	Defective Device	21/01/2023	21/01/2023
									A070504	Premature Discharge of Battery	21/01/2023	21/01/2023
					CT2727	Active implant devices/ancillaries	E2403	No Clinical Signs, Symptoms or Conditions	A0203	Defective Device	21/01/2023	21/01/2023
									A070504	Premature Discharge of Battery	21/01/2023	21/01/2023
							F08	Hospitalization or Prolonged Hospitalization	A0203	Defective Device	21/01/2023	21/01/2023
									A070504	Premature Discharge of Battery	21/01/2023	21/01/2023
							F1905	Device Revision or Replacement	A0203	Defective Device	21/01/2023	21/01/2023
									A070504	Premature Discharge of Battery	21/01/2023	21/01/2023
					CT2750	Implantable cardiac electrical stimulators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A0203	Defective Device	21/01/2023	21/01/2023
									A070504	Premature Discharge of Battery	21/01/2023	21/01/2023
							F08	Hospitalization or Prolonged Hospitalization	A0203	Defective Device	21/01/2023	21/01/2023
									A070504	Premature Discharge of Battery	21/01/2023	21/01/2023
							F1905	Device Revision or Replacement	A0203	Defective Device	21/01/2023	21/01/2023
									A070504	Premature Discharge of Battery	21/01/2023	21/01/2023
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A0203	Defective Device	21/01/2023	21/01/2023
									A070504	Premature Discharge of Battery	21/01/2023	21/01/2023
							F08	Hospitalization or Prolonged Hospitalization	A0203	Defective Device	21/01/2023	21/01/2023
									A070504	Premature Discharge of Battery	21/01/2023	21/01/2023
							F1905	Device Revision or Replacement	A0203	Defective Device	21/01/2023	21/01/2023
									A070504	Premature Discharge of Battery	21/01/2023	21/01/2023
					CT2985	Active	E2403	No Clinical Signs, Symptoms or Conditions	A0203	Defective Device	21/01/2023	21/01/2023
									A070504	Premature Discharge of Battery	21/01/2023	21/01/2023
							F08	Hospitalization or Prolonged Hospitalization	A0203	Defective Device	21/01/2023	21/01/2023
									A070504	Premature Discharge of Battery	21/01/2023	21/01/2023
							F1905	Device Revision or Replacement	A0203	Defective Device	21/01/2023	21/01/2023
									A070504	Premature Discharge of Battery	21/01/2023	21/01/2023
					CT3001	Implantable metal	E2403	No Clinical Signs, Symptoms or Conditions	A0203	Defective Device	21/01/2023	21/01/2023
									A070504	Premature Discharge of Battery	21/01/2023	21/01/2023
							F08	Hospitalization or Prolonged Hospitalization	A0203	Defective Device	21/01/2023	21/01/2023
									A070504	Premature Discharge of Battery	21/01/2023	21/01/2023
							F1905	Device Revision or Replacement	A0203	Defective Device	21/01/2023	21/01/2023
									A070504	Premature Discharge of Battery	21/01/2023	21/01/2023
					CT3002	Implantable synthetic polymer	E2403	No Clinical Signs, Symptoms or Conditions	A0203	Defective Device	21/01/2023	21/01/2023
									A070504	Premature Discharge of Battery	21/01/2023	21/01/2023

							F08	Hospitalization or Prolonged Hospitalization	A0203	Defective Device	21/01/2023	21/01/2023
									A070504	Premature Discharge of Battery	21/01/2023	21/01/2023
							F1905	Device Revision or Replacement	A0203	Defective Device	21/01/2023	21/01/2023
									A070504	Premature Discharge of Battery	21/01/2023	21/01/2023
					CT3006	Implantable inorganic material	E2403	No Clinical Signs, Symptoms or Conditions	A0203	Defective Device	21/01/2023	21/01/2023
									A070504	Premature Discharge of Battery	21/01/2023	21/01/2023
							F08	Hospitalization or Prolonged Hospitalization	A0203	Defective Device	21/01/2023	21/01/2023
									A070504	Premature Discharge of Battery	21/01/2023	21/01/2023
							F1905	Device Revision or Replacement	A0203	Defective Device	21/01/2023	21/01/2023
									A070504	Premature Discharge of Battery	21/01/2023	21/01/2023
					CT314	Electrically powered	E2403	No Clinical Signs, Symptoms or Conditions	A0203	Defective Device	21/01/2023	21/01/2023
									A070504	Premature Discharge of Battery	21/01/2023	21/01/2023
							F08	Hospitalization or Prolonged Hospitalization	A0203	Defective Device	21/01/2023	21/01/2023
									A070504	Premature Discharge of Battery	21/01/2023	21/01/2023
							F1905	Device Revision or Replacement	A0203	Defective Device	21/01/2023	21/01/2023
									A070504	Premature Discharge of Battery	21/01/2023	21/01/2023
					CT321	Long-term surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A0203	Defective Device	21/01/2023	21/01/2023
									A070504	Premature Discharge of Battery	21/01/2023	21/01/2023
							F08	Hospitalization or Prolonged Hospitalization	A0203	Defective Device	21/01/2023	21/01/2023
									A070504	Premature Discharge of Battery	21/01/2023	21/01/2023
							F1905	Device Revision or Replacement	A0203	Defective Device	21/01/2023	21/01/2023
									A070504	Premature Discharge of Battery	21/01/2023	21/01/2023
					CT449	Generators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A0203	Defective Device	21/01/2023	21/01/2023
									A070504	Premature Discharge of Battery	21/01/2023	21/01/2023
							F08	Hospitalization or Prolonged Hospitalization	A0203	Defective Device	21/01/2023	21/01/2023
									A070504	Premature Discharge of Battery	21/01/2023	21/01/2023
							F1905	Device Revision or Replacement	A0203	Defective Device	21/01/2023	21/01/2023
									A070504	Premature Discharge of Battery	21/01/2023	21/01/2023
					CT981	Single use	E2403	No Clinical Signs, Symptoms or Conditions	A0203	Defective Device	21/01/2023	21/01/2023
									A070504	Premature Discharge of Battery	21/01/2023	21/01/2023
							F08	Hospitalization or Prolonged Hospitalization	A0203	Defective Device	21/01/2023	21/01/2023
									A070504	Premature Discharge of Battery	21/01/2023	21/01/2023
							F1905	Device Revision or Replacement	A0203	Defective Device	21/01/2023	21/01/2023
									A070504	Premature Discharge of Battery	21/01/2023	21/01/2023
					CT983	Surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A0203	Defective Device	21/01/2023	21/01/2023
									A070504	Premature Discharge of Battery	21/01/2023	21/01/2023
							F08	Hospitalization or Prolonged Hospitalization	A0203	Defective Device	21/01/2023	21/01/2023
									A070504	Premature Discharge of Battery	21/01/2023	21/01/2023
							F1905	Device Revision or Replacement	A0203	Defective Device	21/01/2023	21/01/2023
									A070504	Premature Discharge of Battery	21/01/2023	21/01/2023
14554	2023	All other reportable incidents	47267	Single-chamber implantable pacemaker, rate-responsive	CT141	Electrodes/Leads and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	05/10/2022	05/10/2022
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	05/10/2022	05/10/2022
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	05/10/2022	05/10/2022
					CT145	Cardiovascular devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	05/10/2022	05/10/2022
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	05/10/2022	05/10/2022
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	05/10/2022	05/10/2022
					CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	05/10/2022	05/10/2022
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	05/10/2022	05/10/2022
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	05/10/2022	05/10/2022
					CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	05/10/2022	05/10/2022
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	05/10/2022	05/10/2022
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	05/10/2022	05/10/2022
					CT233	Surgical	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	05/10/2022	05/10/2022
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	05/10/2022	05/10/2022
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	05/10/2022	05/10/2022
					CT2406	Implantable	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	05/10/2022	05/10/2022
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	05/10/2022	05/10/2022
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	05/10/2022	05/10/2022
					CT2407	Active implantable	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	05/10/2022	05/10/2022
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	05/10/2022	05/10/2022
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	05/10/2022	05/10/2022
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	05/10/2022	05/10/2022
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	05/10/2022	05/10/2022
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	05/10/2022	05/10/2022
					CT2727	Active implant devices/ancillaries	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	05/10/2022	05/10/2022
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	05/10/2022	05/10/2022
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	05/10/2022	05/10/2022
					CT2750	Implantable cardiac electrical stimulators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	05/10/2022	05/10/2022
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	05/10/2022	05/10/2022
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	05/10/2022	05/10/2022
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	05/10/2022	05/10/2022
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	05/10/2022	05/10/2022
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	05/10/2022	05/10/2022
					CT2985	Active	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	05/10/2022	05/10/2022
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	05/10/2022	05/10/2022
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	05/10/2022	05/10/2022
					CT3001	Implantable metal	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	05/10/2022	05/10/2022
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	05/10/2022	05/10/2022
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	05/10/2022	05/10/2022
					CT3002	Implantable synthetic polymer	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	05/10/2022	05/10/2022
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	05/10/2022	05/10/2022
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	05/10/2022	05/10/2022
					CT3006	Implantable inorganic material	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	05/10/2022	05/10/2022
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	05/10/2022	05/10/2022
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	05/10/2022	05/10/2022
					CT314	Electrically powered	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	05/10/2022	05/10/2022
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	05/10/2022	05/10/2022
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	05/10/2022	05/10/2022

					CT321	Long-term surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	05/10/2022	05/10/2022
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	05/10/2022	05/10/2022
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	05/10/2022	05/10/2022
					CT449	Generators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	05/10/2022	05/10/2022
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	05/10/2022	05/10/2022
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	05/10/2022	05/10/2022
					CT981	Single use	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	05/10/2022	05/10/2022
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	05/10/2022	05/10/2022
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	05/10/2022	05/10/2022
					CT983	Surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	05/10/2022	05/10/2022
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	05/10/2022	05/10/2022
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	05/10/2022	05/10/2022
14555	2023	All other reportable incidents	35852	Single-chamber implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
14556	2023	All other reportable incidents	47265	Dual-chamber implantable pacemaker, rate-responsive	CT1685	Cardiac pacemakers	E060101	Asystole	A070101	Failure to Capture		
									A070909	Oversensing		
									A0712	Pacing Problem		
									A090801	Signal Artifact/Noise		
							F1001	Absence of Treatment	A070101	Failure to Capture		
									A070909	Oversensing		
									A0712	Pacing Problem		
									A090801	Signal Artifact/Noise		
					CT257	Implantable cardiac pacemakers	E060101	Asystole	A070101	Failure to Capture		
									A070909	Oversensing		
									A0712	Pacing Problem		
									A090801	Signal Artifact/Noise		
							F1001	Absence of Treatment	A070101	Failure to Capture		
									A070909	Oversensing		
									A0712	Pacing Problem		
									A090801	Signal Artifact/Noise		
					CT2866	Implantable cardiac electrical stimulators	E060101	Asystole	A070101	Failure to Capture		
									A070909	Oversensing		
									A0712	Pacing Problem		
									A090801	Signal Artifact/Noise		
							F1001	Absence of Treatment	A070101	Failure to Capture		
									A070909	Oversensing		
									A0712	Pacing Problem		
									A090801	Signal Artifact/Noise		
14557	2023	All other reportable incidents	47265	Dual-chamber implantable pacemaker, rate-responsive	CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A0723	Unexpected Mode Switch	30/11/2023	30/11/2023
							F1901	Additional Surgery	A0723	Unexpected Mode Switch	30/11/2023	30/11/2023
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A0723	Unexpected Mode Switch	30/11/2023	30/11/2023
							F1901	Additional Surgery	A0723	Unexpected Mode Switch	30/11/2023	30/11/2023
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A0723	Unexpected Mode Switch	30/11/2023	30/11/2023
							F1901	Additional Surgery	A0723	Unexpected Mode Switch	30/11/2023	30/11/2023
14558	2023	All other reportable incidents	35852	Single-chamber implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	10/01/2024	10/01/2024
									A0908	Incorrect, Inadequate or Imprecise Result or Readings	10/01/2024	10/01/2024
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	10/01/2024	10/01/2024
									A0908	Incorrect, Inadequate or Imprecise Result or Readings	10/01/2024	10/01/2024
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	10/01/2024	10/01/2024
									A0908	Incorrect, Inadequate or Imprecise Result or Readings	10/01/2024	10/01/2024
14559	2023	All other reportable incidents	35852	Single-chamber implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A0709	Device Sensing Problem		
									A070909	Oversensing		
							F26	No Health Consequences or Impact	A0709	Device Sensing Problem		
									A070909	Oversensing		
14560	2023	All other reportable incidents	47265	Dual-chamber implantable pacemaker, rate-responsive	CT141	Electrodes/Leads and associated devices	E0601	Arrhythmia	A071206	Pacing Asynchronously		
							F23	Unexpected Medical Intervention	A071206	Pacing Asynchronously		
					CT145	Cardiovascular devices	E0601	Arrhythmia	A071206	Pacing Asynchronously		
							F23	Unexpected Medical Intervention	A071206	Pacing Asynchronously		
					CT1684	Cardiac pacemakers and associated devices	E0601	Arrhythmia	A071206	Pacing Asynchronously		
							F23	Unexpected Medical Intervention	A071206	Pacing Asynchronously		
					CT1685	Cardiac pacemakers	E0601	Arrhythmia	A071206	Pacing Asynchronously		
							F23	Unexpected Medical Intervention	A071206	Pacing Asynchronously		
					CT233	Surgical	E0601	Arrhythmia	A071206	Pacing Asynchronously		
							F23	Unexpected Medical Intervention	A071206	Pacing Asynchronously		
					CT2406	Implantable	E0601	Arrhythmia	A071206	Pacing Asynchronously		
							F23	Unexpected Medical Intervention	A071206	Pacing Asynchronously		
					CT2407	Active implantable	E0601	Arrhythmia	A071206	Pacing Asynchronously		
							F23	Unexpected Medical Intervention	A071206	Pacing Asynchronously		
					CT257	Implantable cardiac pacemakers	E0601	Arrhythmia	A071206	Pacing Asynchronously		
							F23	Unexpected Medical Intervention	A071206	Pacing Asynchronously		
					CT2727	Active implant devices/ancillaries	E0601	Arrhythmia	A071206	Pacing Asynchronously		
							F23	Unexpected Medical Intervention	A071206	Pacing Asynchronously		
					CT2750	Implantable cardiac electrical stimulators and associated devices	E0601	Arrhythmia	A071206	Pacing Asynchronously		
							F23	Unexpected Medical Intervention	A071206	Pacing Asynchronously		
					CT2866	Implantable cardiac electrical stimulators	E0601	Arrhythmia	A071206	Pacing Asynchronously		
							F23	Unexpected Medical Intervention	A071206	Pacing Asynchronously		
					CT2985	Active	E0601	Arrhythmia	A071206	Pacing Asynchronously		
							F23	Unexpected Medical Intervention	A071206	Pacing Asynchronously		
					CT3001	Implantable metal	E0601	Arrhythmia	A071206	Pacing Asynchronously		
							F23	Unexpected Medical Intervention	A071206	Pacing Asynchronously		
					CT3002	Implantable synthetic polymer	E0601	Arrhythmia	A071206	Pacing Asynchronously		
							F23	Unexpected Medical Intervention	A071206	Pacing Asynchronously		
					CT3006	Implantable inorganic material	E0601	Arrhythmia	A071206	Pacing Asynchronously		

							F23	Unexpected Medical Intervention	A071206	Pacing Asynchronously		
					CT314	Electrically powered	E0601	Arrhythmia	A071206	Pacing Asynchronously		
							F23	Unexpected Medical Intervention	A071206	Pacing Asynchronously		
					CT321	Long-term surgical invasive	E0601	Arrhythmia	A071206	Pacing Asynchronously		
							F23	Unexpected Medical Intervention	A071206	Pacing Asynchronously		
					CT449	Generators and associated devices	E0601	Arrhythmia	A071206	Pacing Asynchronously		
							F23	Unexpected Medical Intervention	A071206	Pacing Asynchronously		
					CT981	Single use	E0601	Arrhythmia	A071206	Pacing Asynchronously		
							F23	Unexpected Medical Intervention	A071206	Pacing Asynchronously		
					CT983	Surgical invasive	E0601	Arrhythmia	A071206	Pacing Asynchronously		
							F23	Unexpected Medical Intervention	A071206	Pacing Asynchronously		
14561	2023	All other reportable incidents	17882	Physiologic-monitoring defibrillation system	CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070803	Failure to Power Up		
							F27	Problem Identified During Non-Clinical Procedure	A070803	Failure to Power Up		
14562	2023	All other reportable incidents	47265	Dual-chamber implantable pacemaker, rate-responsive	CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A070909	Oversensing		
							F26	No Health Consequences or Impact	A070909	Oversensing		
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A070909	Oversensing		
							F26	No Health Consequences or Impact	A070909	Oversensing		
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A070909	Oversensing		
							F26	No Health Consequences or Impact	A070909	Oversensing		
14563	2023		47265	Dual-chamber implantable pacemaker, rate-responsive	CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	21/11/2023	21/11/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	21/11/2023	21/11/2023
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	21/11/2023	21/11/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	21/11/2023	21/11/2023
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	21/11/2023	21/11/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	21/11/2023	21/11/2023
14564	2023	All other reportable incidents	47270	Cardiac resynchronization therapy implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E060109	Tachycardia	A0709	Device Sensing Problem		
									A071301	Failure to Deliver Shock/Stimulation		
							F02	Death	A0709	Device Sensing Problem		
									A071301	Failure to Deliver Shock/Stimulation		
							F1203	Life Threatening Illness or Injury	A0709	Device Sensing Problem		
									A071301	Failure to Deliver Shock/Stimulation		
14565	2023	All other reportable incidents	17882	Physiologic-monitoring defibrillation system	CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A0719	Unexpected Shutdown		
							F26	No Health Consequences or Impact	A0719	Unexpected Shutdown		
14566	2023	All other reportable incidents	47265	Dual-chamber implantable pacemaker, rate-responsive	CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	06/12/2023	06/12/2023
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	06/12/2023	06/12/2023
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	06/12/2023	06/12/2023
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	06/12/2023	06/12/2023
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	06/12/2023	06/12/2023
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	06/12/2023	06/12/2023
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	06/12/2023	06/12/2023
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	06/12/2023	06/12/2023
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	06/12/2023	06/12/2023
14567	2023	All other reportable incidents			CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070909	Oversensing		
									A1301	Failure to Read Input Signal		
							F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A070909	Oversensing		
									A1301	Failure to Read Input Signal		
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A070909	Oversensing		
									A1301	Failure to Read Input Signal		
							F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A070909	Oversensing		
									A1301	Failure to Read Input Signal		
14568	2023	All other reportable incidents	35852	Single-chamber implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
									A0908	Incorrect, Inadequate or Imprecise Result or Readings		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
									A0908	Incorrect, Inadequate or Imprecise Result or Readings		
14569	2023	All other reportable incidents	35852	Single-chamber implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A0712	Pacing Problem		
							F1001	Absence of Treatment	A0712	Pacing Problem		
14570	2023	All other reportable incidents	47265	Dual-chamber implantable pacemaker, rate-responsive	CT141	Electrodes/Leads and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem	03/03/2025	03/03/2025
									A070504	Premature Discharge of Battery	03/03/2025	03/03/2025
							F08	Hospitalization or Prolonged Hospitalization	A0705	Battery Problem	03/03/2025	03/03/2025
									A070504	Premature Discharge of Battery	03/03/2025	03/03/2025
							F1905	Device Revision or Replacement	A0705	Battery Problem	03/03/2025	03/03/2025
									A070504	Premature Discharge of Battery	03/03/2025	03/03/2025
					CT145	Cardiovascular devices	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem	03/03/2025	03/03/2025
									A070504	Premature Discharge of Battery	03/03/2025	03/03/2025
							F08	Hospitalization or Prolonged Hospitalization	A0705	Battery Problem	03/03/2025	03/03/2025
									A070504	Premature Discharge of Battery	03/03/2025	03/03/2025
							F1905	Device Revision or Replacement	A0705	Battery Problem	03/03/2025	03/03/2025
									A070504	Premature Discharge of Battery	03/03/2025	03/03/2025
					CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem	03/03/2025	03/03/2025
									A070504	Premature Discharge of Battery	03/03/2025	03/03/2025
							F08	Hospitalization or Prolonged Hospitalization	A0705	Battery Problem	03/03/2025	03/03/2025
									A070504	Premature Discharge of Battery	03/03/2025	03/03/2025
							F1905	Device Revision or Replacement	A0705	Battery Problem	03/03/2025	03/03/2025
									A070504	Premature Discharge of Battery	03/03/2025	03/03/2025
					CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem	03/03/2025	03/03/2025
									A070504	Premature Discharge of Battery	03/03/2025	03/03/2025
							F08	Hospitalization or Prolonged Hospitalization	A0705	Battery Problem	03/03/2025	03/03/2025
									A070504	Premature Discharge of Battery	03/03/2025	03/03/2025
							F1905	Device Revision or Replacement	A0705	Battery Problem	03/03/2025	03/03/2025
									A070504	Premature Discharge of Battery	03/03/2025	03/03/2025
					CT233	Surgical	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem	03/03/2025	03/03/2025
									A070504	Premature Discharge of Battery	03/03/2025	03/03/2025
							F08	Hospitalization or Prolonged Hospitalization	A0705	Battery Problem	03/03/2025	03/03/2025
									A070504	Premature Discharge of Battery	03/03/2025	03/03/2025
							F1905	Device Revision or Replacement	A0705	Battery Problem	03/03/2025	03/03/2025
									A070504	Premature Discharge of Battery	03/03/2025	03/03/2025

CT2406	Implantable	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem	03/03/2025	03/03/2025
				A070504	Premature Discharge of Battery	03/03/2025	03/03/2025
		F08	Hospitalization or Prolonged Hospitalization	A0705	Battery Problem	03/03/2025	03/03/2025
				A070504	Premature Discharge of Battery	03/03/2025	03/03/2025
		F1905	Device Revision or Replacement	A0705	Battery Problem	03/03/2025	03/03/2025
				A070504	Premature Discharge of Battery	03/03/2025	03/03/2025
CT2407	Active implantable	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem	03/03/2025	03/03/2025
				A070504	Premature Discharge of Battery	03/03/2025	03/03/2025
		F08	Hospitalization or Prolonged Hospitalization	A0705	Battery Problem	03/03/2025	03/03/2025
				A070504	Premature Discharge of Battery	03/03/2025	03/03/2025
		F1905	Device Revision or Replacement	A0705	Battery Problem	03/03/2025	03/03/2025
				A070504	Premature Discharge of Battery	03/03/2025	03/03/2025
CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem	03/03/2025	03/03/2025
				A070504	Premature Discharge of Battery	03/03/2025	03/03/2025
		F08	Hospitalization or Prolonged Hospitalization	A0705	Battery Problem	03/03/2025	03/03/2025
				A070504	Premature Discharge of Battery	03/03/2025	03/03/2025
		F1905	Device Revision or Replacement	A0705	Battery Problem	03/03/2025	03/03/2025
				A070504	Premature Discharge of Battery	03/03/2025	03/03/2025
CT2727	Active implant devices/ancillaries	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem	03/03/2025	03/03/2025
				A070504	Premature Discharge of Battery	03/03/2025	03/03/2025
		F08	Hospitalization or Prolonged Hospitalization	A0705	Battery Problem	03/03/2025	03/03/2025
				A070504	Premature Discharge of Battery	03/03/2025	03/03/2025
		F1905	Device Revision or Replacement	A0705	Battery Problem	03/03/2025	03/03/2025
				A070504	Premature Discharge of Battery	03/03/2025	03/03/2025
CT2750	Implantable cardiac electrical stimulators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem	03/03/2025	03/03/2025
				A070504	Premature Discharge of Battery	03/03/2025	03/03/2025
		F08	Hospitalization or Prolonged Hospitalization	A0705	Battery Problem	03/03/2025	03/03/2025
				A070504	Premature Discharge of Battery	03/03/2025	03/03/2025
		F1905	Device Revision or Replacement	A0705	Battery Problem	03/03/2025	03/03/2025
				A070504	Premature Discharge of Battery	03/03/2025	03/03/2025
CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem	03/03/2025	03/03/2025
				A070504	Premature Discharge of Battery	03/03/2025	03/03/2025
		F08	Hospitalization or Prolonged Hospitalization	A0705	Battery Problem	03/03/2025	03/03/2025
				A070504	Premature Discharge of Battery	03/03/2025	03/03/2025
		F1905	Device Revision or Replacement	A0705	Battery Problem	03/03/2025	03/03/2025
				A070504	Premature Discharge of Battery	03/03/2025	03/03/2025
CT2985	Active	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem	03/03/2025	03/03/2025
				A070504	Premature Discharge of Battery	03/03/2025	03/03/2025
		F08	Hospitalization or Prolonged Hospitalization	A0705	Battery Problem	03/03/2025	03/03/2025
				A070504	Premature Discharge of Battery	03/03/2025	03/03/2025
		F1905	Device Revision or Replacement	A0705	Battery Problem	03/03/2025	03/03/2025
				A070504	Premature Discharge of Battery	03/03/2025	03/03/2025
CT3001	Implantable metal	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem	03/03/2025	03/03/2025
				A070504	Premature Discharge of Battery	03/03/2025	03/03/2025
		F08	Hospitalization or Prolonged Hospitalization	A0705	Battery Problem	03/03/2025	03/03/2025
				A070504	Premature Discharge of Battery	03/03/2025	03/03/2025
		F1905	Device Revision or Replacement	A0705	Battery Problem	03/03/2025	03/03/2025
				A070504	Premature Discharge of Battery	03/03/2025	03/03/2025
CT3002	Implantable synthetic polymer	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem	03/03/2025	03/03/2025
				A070504	Premature Discharge of Battery	03/03/2025	03/03/2025
		F08	Hospitalization or Prolonged Hospitalization	A0705	Battery Problem	03/03/2025	03/03/2025
				A070504	Premature Discharge of Battery	03/03/2025	03/03/2025
		F1905	Device Revision or Replacement	A0705	Battery Problem	03/03/2025	03/03/2025
				A070504	Premature Discharge of Battery	03/03/2025	03/03/2025
CT3006	Implantable inorganic material	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem	03/03/2025	03/03/2025
				A070504	Premature Discharge of Battery	03/03/2025	03/03/2025
		F08	Hospitalization or Prolonged Hospitalization	A0705	Battery Problem	03/03/2025	03/03/2025
				A070504	Premature Discharge of Battery	03/03/2025	03/03/2025
		F1905	Device Revision or Replacement	A0705	Battery Problem	03/03/2025	03/03/2025
				A070504	Premature Discharge of Battery	03/03/2025	03/03/2025
CT314	Electrically powered	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem	03/03/2025	03/03/2025
				A070504	Premature Discharge of Battery	03/03/2025	03/03/2025
		F08	Hospitalization or Prolonged Hospitalization	A0705	Battery Problem	03/03/2025	03/03/2025
				A070504	Premature Discharge of Battery	03/03/2025	03/03/2025
		F1905	Device Revision or Replacement	A0705	Battery Problem	03/03/2025	03/03/2025
				A070504	Premature Discharge of Battery	03/03/2025	03/03/2025
CT321	Long-term surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem	03/03/2025	03/03/2025
				A070504	Premature Discharge of Battery	03/03/2025	03/03/2025
		F08	Hospitalization or Prolonged Hospitalization	A0705	Battery Problem	03/03/2025	03/03/2025
				A070504	Premature Discharge of Battery	03/03/2025	03/03/2025
		F1905	Device Revision or Replacement	A0705	Battery Problem	03/03/2025	03/03/2025
				A070504	Premature Discharge of Battery	03/03/2025	03/03/2025
CT449	Generators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem	03/03/2025	03/03/2025
				A070504	Premature Discharge of Battery	03/03/2025	03/03/2025
		F08	Hospitalization or Prolonged Hospitalization	A0705	Battery Problem	03/03/2025	03/03/2025
				A070504	Premature Discharge of Battery	03/03/2025	03/03/2025
		F1905	Device Revision or Replacement	A0705	Battery Problem	03/03/2025	03/03/2025
				A070504	Premature Discharge of Battery	03/03/2025	03/03/2025
CT981	Single use	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem	03/03/2025	03/03/2025
				A070504	Premature Discharge of Battery	03/03/2025	03/03/2025
		F08	Hospitalization or Prolonged Hospitalization	A0705	Battery Problem	03/03/2025	03/03/2025
				A070504	Premature Discharge of Battery	03/03/2025	03/03/2025
		F1905	Device Revision or Replacement	A0705	Battery Problem	03/03/2025	03/03/2025
				A070504	Premature Discharge of Battery	03/03/2025	03/03/2025
CT983	Surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem	03/03/2025	03/03/2025
				A070504	Premature Discharge of Battery	03/03/2025	03/03/2025
		F08	Hospitalization or Prolonged Hospitalization	A0705	Battery Problem	03/03/2025	03/03/2025
				A070504	Premature Discharge of Battery	03/03/2025	03/03/2025

							F1905	Device Revision or Replacement	A0705	Battery Problem	03/03/2025	03/03/2025
									A070504	Premature Discharge of Battery	03/03/2025	03/03/2025
14571	2023	All other reportable incidents	47265	Dual-chamber implantable pacemaker, rate-responsive	CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A070101	Failure to Capture		
							F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A070101	Failure to Capture		
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A070101	Failure to Capture		
							F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A070101	Failure to Capture		
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A070101	Failure to Capture		
							F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A070101	Failure to Capture		
14572	2023	All other reportable incidents	47265	Dual-chamber implantable pacemaker, rate-responsive	CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	06/12/2023	06/12/2023
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	06/12/2023	06/12/2023
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	06/12/2023	06/12/2023
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	06/12/2023	06/12/2023
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	06/12/2023	06/12/2023
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	06/12/2023	06/12/2023
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	06/12/2023	06/12/2023
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	06/12/2023	06/12/2023
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	06/12/2023	06/12/2023
14573	2023	All other reportable incidents	47267	Single-chamber implantable pacemaker, rate-responsive	CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	06/12/2023	06/12/2023
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	06/12/2023	06/12/2023
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	06/12/2023	06/12/2023
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	06/12/2023	06/12/2023
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	06/12/2023	06/12/2023
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	06/12/2023	06/12/2023
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	06/12/2023	06/12/2023
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	06/12/2023	06/12/2023
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	06/12/2023	06/12/2023
14574	2023	All other reportable incidents	47265	Dual-chamber implantable pacemaker, rate-responsive	CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	06/12/2023	06/12/2023
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	06/12/2023	06/12/2023
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	06/12/2023	06/12/2023
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	06/12/2023	06/12/2023
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	06/12/2023	06/12/2023
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	06/12/2023	06/12/2023
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	06/12/2023	06/12/2023
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	06/12/2023	06/12/2023
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	06/12/2023	06/12/2023
14575	2023	All other reportable incidents	17882	Physiologic-monitoring defibrillation system	CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A071301	Failure to Deliver Shock/Stimulation		
							F27	Problem Identified During Non-Clinical Procedure	A071301	Failure to Deliver Shock/Stimulation		
14576	2023		47265	Dual-chamber implantable pacemaker, rate-responsive	CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	30/11/2023	30/11/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	30/11/2023	30/11/2023
							F1903	Device Explantation	A24	No Apparent Device Problem	30/11/2023	30/11/2023
							F1905	Device Revision or Replacement	A24	No Apparent Device Problem	30/11/2023	30/11/2023
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	30/11/2023	30/11/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	30/11/2023	30/11/2023
							F1903	Device Explantation	A24	No Apparent Device Problem	30/11/2023	30/11/2023
							F1905	Device Revision or Replacement	A24	No Apparent Device Problem	30/11/2023	30/11/2023
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	30/11/2023	30/11/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	30/11/2023	30/11/2023
							F1903	Device Explantation	A24	No Apparent Device Problem	30/11/2023	30/11/2023
							F1905	Device Revision or Replacement	A24	No Apparent Device Problem	30/11/2023	30/11/2023
14577	2023		47265	Dual-chamber implantable pacemaker, rate-responsive	CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A1204	Failure to Disconnect	30/11/2023	30/11/2023
							F1901	Additional Surgery	A1204	Failure to Disconnect	30/11/2023	30/11/2023
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A1204	Failure to Disconnect	30/11/2023	30/11/2023
							F1901	Additional Surgery	A1204	Failure to Disconnect	30/11/2023	30/11/2023
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A1204	Failure to Disconnect	30/11/2023	30/11/2023
							F1901	Additional Surgery	A1204	Failure to Disconnect	30/11/2023	30/11/2023
14578	2023	All other reportable incidents	47265	Dual-chamber implantable pacemaker, rate-responsive	CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	01/12/2023	01/12/2023
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	01/12/2023	01/12/2023
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	01/12/2023	01/12/2023
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	01/12/2023	01/12/2023
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	01/12/2023	01/12/2023
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	01/12/2023	01/12/2023
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	01/12/2023	01/12/2023
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	01/12/2023	01/12/2023
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	01/12/2023	01/12/2023
14579	2023	All other reportable incidents	47267	Single-chamber implantable pacemaker, rate-responsive	CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	18/12/2023	18/12/2023
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	18/12/2023	18/12/2023
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	18/12/2023	18/12/2023
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	18/12/2023	18/12/2023
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	18/12/2023	18/12/2023
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	18/12/2023	18/12/2023
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	18/12/2023	18/12/2023
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	18/12/2023	18/12/2023
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	18/12/2023	18/12/2023
14580	2023		47265	Dual-chamber implantable pacemaker, rate-responsive	CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	28/11/2023	28/11/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	28/11/2023	28/11/2023
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	28/11/2023	28/11/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	28/11/2023	28/11/2023
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	28/11/2023	28/11/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	28/11/2023	28/11/2023
14581	2023	All other reportable incidents	47265	Dual-chamber implantable pacemaker, rate-responsive	CT1685	Cardiac pacemakers	E060103	Atrial Flutter	A0712	Pacing Problem		
							E060104	Bradycardia	A0712	Pacing Problem		
							E060109	Tachycardia	A0712	Pacing Problem		
							E2321	Low Blood Pressure/ Hypotension	A0712	Pacing Problem		
							F12	Serious Injury/ Illness/ Impairment	A0712	Pacing Problem		
							F23	Unexpected Medical Intervention	A0712	Pacing Problem		
					CT257	Implantable cardiac pacemakers	E060103	Atrial Flutter	A0712	Pacing Problem		
							E060104	Bradycardia	A0712	Pacing Problem		
							E060109	Tachycardia	A0712	Pacing Problem		

							E2321	Low Blood Pressure/ Hypotension	A0712	Pacing Problem		
							F12	Serious Injury/ Illness/ Impairment	A0712	Pacing Problem		
							F23	Unexpected Medical Intervention	A0712	Pacing Problem		
					CT2866	Implantable cardiac electrical stimulators	E060103	Atrial Flutter	A0712	Pacing Problem		
							E060104	Bradycardia	A0712	Pacing Problem		
							E060109	Tachycardia	A0712	Pacing Problem		
							E2321	Low Blood Pressure/ Hypotension	A0712	Pacing Problem		
							F12	Serious Injury/ Illness/ Impairment	A0712	Pacing Problem		
							F23	Unexpected Medical Intervention	A0712	Pacing Problem		
14582	2023	All other reportable incidents	47265	Dual-chamber implantable pacemaker, rate-responsive	CT1685	Cardiac pacemakers	E0112	Dizziness	A0705	Battery Problem	24/11/2023	24/11/2023
									A070909	Oversensing	24/11/2023	24/11/2023
									A0712	Pacing Problem	24/11/2023	24/11/2023
									A0908	Incorrect, Inadequate or Imprecise Result or Readings	24/11/2023	24/11/2023
							F08	Hospitalization or Prolonged Hospitalization	A0705	Battery Problem	24/11/2023	24/11/2023
									A070909	Oversensing	24/11/2023	24/11/2023
									A0712	Pacing Problem	24/11/2023	24/11/2023
									A0908	Incorrect, Inadequate or Imprecise Result or Readings	24/11/2023	24/11/2023
							F1001	Absence of Treatment	A0705	Battery Problem	24/11/2023	24/11/2023
									A070909	Oversensing	24/11/2023	24/11/2023
									A0712	Pacing Problem	24/11/2023	24/11/2023
									A0908	Incorrect, Inadequate or Imprecise Result or Readings	24/11/2023	24/11/2023
							F1905	Device Revision or Replacement	A0705	Battery Problem	24/11/2023	24/11/2023
									A070909	Oversensing	24/11/2023	24/11/2023
									A0712	Pacing Problem	24/11/2023	24/11/2023
									A0908	Incorrect, Inadequate or Imprecise Result or Readings	24/11/2023	24/11/2023
					CT257	Implantable cardiac pacemakers	E0112	Dizziness	A0705	Battery Problem	24/11/2023	24/11/2023
									A070909	Oversensing	24/11/2023	24/11/2023
									A0712	Pacing Problem	24/11/2023	24/11/2023
									A0908	Incorrect, Inadequate or Imprecise Result or Readings	24/11/2023	24/11/2023
							F08	Hospitalization or Prolonged Hospitalization	A0705	Battery Problem	24/11/2023	24/11/2023
									A070909	Oversensing	24/11/2023	24/11/2023
									A0712	Pacing Problem	24/11/2023	24/11/2023
									A0908	Incorrect, Inadequate or Imprecise Result or Readings	24/11/2023	24/11/2023
							F1001	Absence of Treatment	A0705	Battery Problem	24/11/2023	24/11/2023
									A070909	Oversensing	24/11/2023	24/11/2023
									A0712	Pacing Problem	24/11/2023	24/11/2023
									A0908	Incorrect, Inadequate or Imprecise Result or Readings	24/11/2023	24/11/2023
							F1905	Device Revision or Replacement	A0705	Battery Problem	24/11/2023	24/11/2023
									A070909	Oversensing	24/11/2023	24/11/2023
									A0712	Pacing Problem	24/11/2023	24/11/2023
									A0908	Incorrect, Inadequate or Imprecise Result or Readings	24/11/2023	24/11/2023
					CT2866	Implantable cardiac electrical stimulators	E0112	Dizziness	A0705	Battery Problem	24/11/2023	24/11/2023
									A070909	Oversensing	24/11/2023	24/11/2023
									A0712	Pacing Problem	24/11/2023	24/11/2023
									A0908	Incorrect, Inadequate or Imprecise Result or Readings	24/11/2023	24/11/2023
							F08	Hospitalization or Prolonged Hospitalization	A0705	Battery Problem	24/11/2023	24/11/2023
									A070909	Oversensing	24/11/2023	24/11/2023
									A0712	Pacing Problem	24/11/2023	24/11/2023
									A0908	Incorrect, Inadequate or Imprecise Result or Readings	24/11/2023	24/11/2023
							F1001	Absence of Treatment	A0705	Battery Problem	24/11/2023	24/11/2023
									A070909	Oversensing	24/11/2023	24/11/2023
									A0712	Pacing Problem	24/11/2023	24/11/2023
									A0908	Incorrect, Inadequate or Imprecise Result or Readings	24/11/2023	24/11/2023
							F1905	Device Revision or Replacement	A0705	Battery Problem	24/11/2023	24/11/2023
									A070909	Oversensing	24/11/2023	24/11/2023
									A0712	Pacing Problem	24/11/2023	24/11/2023
									A0908	Incorrect, Inadequate or Imprecise Result or Readings	24/11/2023	24/11/2023
14583	2023	All other reportable incidents	47263	Cardiac resynchronization therapy implantable pacemaker	CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A070910	Undersensing	29/08/2023	29/08/2023
									A0722	Impedance Problem	29/08/2023	29/08/2023
							F08	Hospitalization or Prolonged Hospitalization	A070910	Undersensing	29/08/2023	29/08/2023
									A0722	Impedance Problem	29/08/2023	29/08/2023
							F1905	Device Revision or Replacement	A070910	Undersensing	29/08/2023	29/08/2023
									A0722	Impedance Problem	29/08/2023	29/08/2023
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A070910	Undersensing	29/08/2023	29/08/2023
									A0722	Impedance Problem	29/08/2023	29/08/2023
							F08	Hospitalization or Prolonged Hospitalization	A070910	Undersensing	29/08/2023	29/08/2023
									A0722	Impedance Problem	29/08/2023	29/08/2023
							F1905	Device Revision or Replacement	A070910	Undersensing	29/08/2023	29/08/2023
									A0722	Impedance Problem	29/08/2023	29/08/2023
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A070910	Undersensing	29/08/2023	29/08/2023
									A0722	Impedance Problem	29/08/2023	29/08/2023
							F08	Hospitalization or Prolonged Hospitalization	A070910	Undersensing	29/08/2023	29/08/2023
									A0722	Impedance Problem	29/08/2023	29/08/2023
							F1905	Device Revision or Replacement	A070910	Undersensing	29/08/2023	29/08/2023
									A0722	Impedance Problem	29/08/2023	29/08/2023
14584	2023	All other reportable incidents	35852	Single-chamber implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E2104	Electric Shock	A071302	Inappropriate/Inadequate Shock/Stimulation	28/11/2023	28/11/2023
									A0908	Incorrect, Inadequate or Imprecise Result or Readings	28/11/2023	28/11/2023
							F08	Hospitalization or Prolonged Hospitalization	A071302	Inappropriate/Inadequate Shock/Stimulation	28/11/2023	28/11/2023

									A0908	Incorrect, Inadequate or Imprecise Result or Readings	28/11/2023	28/11/2023
							F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A071302	Inappropriate/Inadequate Shock/Stimulation	28/11/2023	28/11/2023
									A0908	Incorrect, Inadequate or Imprecise Result or Readings	28/11/2023	28/11/2023
							F1905	Device Revision or Replacement	A071302	Inappropriate/Inadequate Shock/Stimulation	28/11/2023	28/11/2023
									A0908	Incorrect, Inadequate or Imprecise Result or Readings	28/11/2023	28/11/2023
							F23	Unexpected Medical Intervention	A071302	Inappropriate/Inadequate Shock/Stimulation	28/11/2023	28/11/2023
									A0908	Incorrect, Inadequate or Imprecise Result or Readings	28/11/2023	28/11/2023
14585	2023	All other reportable incidents	47265	Dual-chamber implantable pacemaker, rate-responsive	CT1685	Cardiac pacemakers	E0601	Arrhythmia	A070504	Premature Discharge of Battery	07/12/2023	07/12/2023
									A070909	Oversensing	07/12/2023	07/12/2023
									A071206	Pacing Asynchronously	07/12/2023	07/12/2023
									A0723	Unexpected Mode Switch	07/12/2023	07/12/2023
							F1901	Additional Surgery	A070504	Premature Discharge of Battery	07/12/2023	07/12/2023
									A070909	Oversensing	07/12/2023	07/12/2023
									A071206	Pacing Asynchronously	07/12/2023	07/12/2023
									A0723	Unexpected Mode Switch	07/12/2023	07/12/2023
							F1903	Device Explantation	A070504	Premature Discharge of Battery	07/12/2023	07/12/2023
									A070909	Oversensing	07/12/2023	07/12/2023
									A071206	Pacing Asynchronously	07/12/2023	07/12/2023
									A0723	Unexpected Mode Switch	07/12/2023	07/12/2023
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	07/12/2023	07/12/2023
									A070909	Oversensing	07/12/2023	07/12/2023
									A071206	Pacing Asynchronously	07/12/2023	07/12/2023
									A0723	Unexpected Mode Switch	07/12/2023	07/12/2023
					CT257	Implantable cardiac pacemakers	E0601	Arrhythmia	A070504	Premature Discharge of Battery	07/12/2023	07/12/2023
									A070909	Oversensing	07/12/2023	07/12/2023
									A071206	Pacing Asynchronously	07/12/2023	07/12/2023
									A0723	Unexpected Mode Switch	07/12/2023	07/12/2023
							F1901	Additional Surgery	A070504	Premature Discharge of Battery	07/12/2023	07/12/2023
									A070909	Oversensing	07/12/2023	07/12/2023
									A071206	Pacing Asynchronously	07/12/2023	07/12/2023
									A0723	Unexpected Mode Switch	07/12/2023	07/12/2023
							F1903	Device Explantation	A070504	Premature Discharge of Battery	07/12/2023	07/12/2023
									A070909	Oversensing	07/12/2023	07/12/2023
									A071206	Pacing Asynchronously	07/12/2023	07/12/2023
									A0723	Unexpected Mode Switch	07/12/2023	07/12/2023
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	07/12/2023	07/12/2023
									A070909	Oversensing	07/12/2023	07/12/2023
									A071206	Pacing Asynchronously	07/12/2023	07/12/2023
									A0723	Unexpected Mode Switch	07/12/2023	07/12/2023
					CT2866	Implantable cardiac electrical stimulators	E0601	Arrhythmia	A070504	Premature Discharge of Battery	07/12/2023	07/12/2023
									A070909	Oversensing	07/12/2023	07/12/2023
									A071206	Pacing Asynchronously	07/12/2023	07/12/2023
									A0723	Unexpected Mode Switch	07/12/2023	07/12/2023
							F1901	Additional Surgery	A070504	Premature Discharge of Battery	07/12/2023	07/12/2023
									A070909	Oversensing	07/12/2023	07/12/2023
									A071206	Pacing Asynchronously	07/12/2023	07/12/2023
									A0723	Unexpected Mode Switch	07/12/2023	07/12/2023
							F1903	Device Explantation	A070504	Premature Discharge of Battery	07/12/2023	07/12/2023
									A070909	Oversensing	07/12/2023	07/12/2023
									A071206	Pacing Asynchronously	07/12/2023	07/12/2023
									A0723	Unexpected Mode Switch	07/12/2023	07/12/2023
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	07/12/2023	07/12/2023
									A070909	Oversensing	07/12/2023	07/12/2023
									A071206	Pacing Asynchronously	07/12/2023	07/12/2023
									A0723	Unexpected Mode Switch	07/12/2023	07/12/2023
14586	2023		47265	Dual-chamber implantable pacemaker, rate-responsive	CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	21/11/2023	21/11/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	21/11/2023	21/11/2023
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	21/11/2023	21/11/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	21/11/2023	21/11/2023
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	21/11/2023	21/11/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	21/11/2023	21/11/2023
14587	2023		35223	Endocardial/interventricular septal pacing lead	CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A04	Material Integrity Problem		
									A070101	Failure to Capture		
									A0709	Device Sensing Problem		
									A0722	Impedance Problem		
							F26	No Health Consequences or Impact	A04	Material Integrity Problem		
									A070101	Failure to Capture		
									A0709	Device Sensing Problem		
									A0722	Impedance Problem		
14588	2023	All other reportable incidents	47267	Single-chamber implantable pacemaker, rate-responsive	CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
14589	2023	All other reportable incidents	17882	Physiologic-monitoring defibrillation system	CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem		
							F27	Problem Identified During Non-Clinical Procedure	A0705	Battery Problem		
14590	2023	All other reportable incidents	63188	Left-ventricular wireless pacing system sensor/transmitter	CT145	Cardiovascular devices	E2403	No Clinical Signs, Symptoms or Conditions	A0701	Capturing Problem		
					CT1577	Wireless monitoring system transmitters	E2403	No Clinical Signs, Symptoms or Conditions	A0701	Capturing Problem		

					CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A0701	Capturing Problem		
					CT2727	Active implant devices/ancillaries	E2403	No Clinical Signs, Symptoms or Conditions	A0701	Capturing Problem		
					CT2750	Implantable cardiac electrical stimulators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A0701	Capturing Problem		
					CT2885	Physiological monitoring/measuring system transmitters/receivers/transceivers	E2403	No Clinical Signs, Symptoms or Conditions	A0701	Capturing Problem		
14591	2023	All other reportable incidents	47263	Cardiac resynchronization therapy implantable pacemaker	CT1685	Cardiac pacemakers	E011903	Syncope/Fainting	A070909	Oversensing	17/08/2023	17/08/2023
									A0712	Pacing Problem	17/08/2023	17/08/2023
									A0908	Incorrect, Inadequate or Imprecise Result or Readings	17/08/2023	17/08/2023
							E060101	Asystole	A070909	Oversensing	17/08/2023	17/08/2023
									A0712	Pacing Problem	17/08/2023	17/08/2023
									A0908	Incorrect, Inadequate or Imprecise Result or Readings	17/08/2023	17/08/2023
							F08	Hospitalization or Prolonged Hospitalization	A070909	Oversensing	17/08/2023	17/08/2023
									A0712	Pacing Problem	17/08/2023	17/08/2023
									A0908	Incorrect, Inadequate or Imprecise Result or Readings	17/08/2023	17/08/2023
							F1203	Life Threatening Illness or Injury	A070909	Oversensing	17/08/2023	17/08/2023
									A0712	Pacing Problem	17/08/2023	17/08/2023
									A0908	Incorrect, Inadequate or Imprecise Result or Readings	17/08/2023	17/08/2023
							F1905	Device Revision or Replacement	A070909	Oversensing	17/08/2023	17/08/2023
									A0712	Pacing Problem	17/08/2023	17/08/2023
									A0908	Incorrect, Inadequate or Imprecise Result or Readings	17/08/2023	17/08/2023
							F1907	More Complex Surgery	A070909	Oversensing	17/08/2023	17/08/2023
									A0712	Pacing Problem	17/08/2023	17/08/2023
									A0908	Incorrect, Inadequate or Imprecise Result or Readings	17/08/2023	17/08/2023
					CT257	Implantable cardiac pacemakers	E011903	Syncope/Fainting	A070909	Oversensing	17/08/2023	17/08/2023
									A0712	Pacing Problem	17/08/2023	17/08/2023
									A0908	Incorrect, Inadequate or Imprecise Result or Readings	17/08/2023	17/08/2023
							E060101	Asystole	A070909	Oversensing	17/08/2023	17/08/2023
									A0712	Pacing Problem	17/08/2023	17/08/2023
									A0908	Incorrect, Inadequate or Imprecise Result or Readings	17/08/2023	17/08/2023
							F08	Hospitalization or Prolonged Hospitalization	A070909	Oversensing	17/08/2023	17/08/2023
									A0712	Pacing Problem	17/08/2023	17/08/2023
									A0908	Incorrect, Inadequate or Imprecise Result or Readings	17/08/2023	17/08/2023
							F1203	Life Threatening Illness or Injury	A070909	Oversensing	17/08/2023	17/08/2023
									A0712	Pacing Problem	17/08/2023	17/08/2023
									A0908	Incorrect, Inadequate or Imprecise Result or Readings	17/08/2023	17/08/2023
							F1905	Device Revision or Replacement	A070909	Oversensing	17/08/2023	17/08/2023
									A0712	Pacing Problem	17/08/2023	17/08/2023
									A0908	Incorrect, Inadequate or Imprecise Result or Readings	17/08/2023	17/08/2023
							F1907	More Complex Surgery	A070909	Oversensing	17/08/2023	17/08/2023
									A0712	Pacing Problem	17/08/2023	17/08/2023
									A0908	Incorrect, Inadequate or Imprecise Result or Readings	17/08/2023	17/08/2023
					CT2866	Implantable cardiac electrical stimulators	E011903	Syncope/Fainting	A070909	Oversensing	17/08/2023	17/08/2023
									A0712	Pacing Problem	17/08/2023	17/08/2023
									A0908	Incorrect, Inadequate or Imprecise Result or Readings	17/08/2023	17/08/2023
							E060101	Asystole	A070909	Oversensing	17/08/2023	17/08/2023
									A0712	Pacing Problem	17/08/2023	17/08/2023
									A0908	Incorrect, Inadequate or Imprecise Result or Readings	17/08/2023	17/08/2023
							F08	Hospitalization or Prolonged Hospitalization	A070909	Oversensing	17/08/2023	17/08/2023
									A0712	Pacing Problem	17/08/2023	17/08/2023
									A0908	Incorrect, Inadequate or Imprecise Result or Readings	17/08/2023	17/08/2023
							F1203	Life Threatening Illness or Injury	A070909	Oversensing	17/08/2023	17/08/2023
									A0712	Pacing Problem	17/08/2023	17/08/2023
									A0908	Incorrect, Inadequate or Imprecise Result or Readings	17/08/2023	17/08/2023
							F1905	Device Revision or Replacement	A070909	Oversensing	17/08/2023	17/08/2023
									A0712	Pacing Problem	17/08/2023	17/08/2023
									A0908	Incorrect, Inadequate or Imprecise Result or Readings	17/08/2023	17/08/2023
							F1907	More Complex Surgery	A070909	Oversensing	17/08/2023	17/08/2023
									A0712	Pacing Problem	17/08/2023	17/08/2023
									A0908	Incorrect, Inadequate or Imprecise Result or Readings	17/08/2023	17/08/2023
14592	2023		47265	Dual-chamber implantable pacemaker, rate-responsive	CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	18/08/2023	18/08/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	18/08/2023	18/08/2023
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	18/08/2023	18/08/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	18/08/2023	18/08/2023
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	18/08/2023	18/08/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	18/08/2023	18/08/2023
14593	2023	All other reportable incidents	60789	Intracardiac pacemaker	CT1685	Cardiac pacemakers	E0601	Arrhythmia	A26	Insufficient Device Problem Information		
							F12	Serious Injury/ Illness/ Impairment	A26	Insufficient Device Problem Information		
							F23	Unexpected Medical Intervention	A26	Insufficient Device Problem Information		
					CT257	Implantable cardiac pacemakers	E0601	Arrhythmia	A26	Insufficient Device Problem Information		
							F12	Serious Injury/ Illness/ Impairment	A26	Insufficient Device Problem Information		
							F23	Unexpected Medical Intervention	A26	Insufficient Device Problem Information		
					CT2818	Catheterization/Cannulation/Fluid path devices	E0601	Arrhythmia	A26	Insufficient Device Problem Information		
							F12	Serious Injury/ Illness/ Impairment	A26	Insufficient Device Problem Information		

							F23	Unexpected Medical Intervention	A26	Insufficient Device Problem Information		
					CT2866	Implantable cardiac electrical stimulators	E0601	Arrhythmia	A26	Insufficient Device Problem Information		
							F12	Serious Injury/ Illness/ Impairment	A26	Insufficient Device Problem Information		
							F23	Unexpected Medical Intervention	A26	Insufficient Device Problem Information		
14594	2023		47265	Dual-chamber implantable pacemaker, rate-responsive	CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	14/11/2023	14/11/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	14/11/2023	14/11/2023
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	14/11/2023	14/11/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	14/11/2023	14/11/2023
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	14/11/2023	14/11/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	14/11/2023	14/11/2023
14595	2023	All other reportable incidents	17882	Physiologic-monitoring defibrillation system	CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A090101	Inappropriate Audible Prompt/Feedback		
							F26	No Health Consequences or Impact	A090101	Inappropriate Audible Prompt/Feedback		
14596	2023	All other reportable incidents	47265	Dual-chamber implantable pacemaker, rate-responsive	CT141	Electrodes/Leads and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem		
							F08	Hospitalization or Prolonged Hospitalization	A0705	Battery Problem		
							F12	Serious Injury/ Illness/ Impairment	A0705	Battery Problem		
							F1905	Device Revision or Replacement	A0705	Battery Problem		
							F23	Unexpected Medical Intervention	A0705	Battery Problem		
					CT145	Cardiovascular devices	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem		
							F08	Hospitalization or Prolonged Hospitalization	A0705	Battery Problem		
							F12	Serious Injury/ Illness/ Impairment	A0705	Battery Problem		
							F1905	Device Revision or Replacement	A0705	Battery Problem		
							F23	Unexpected Medical Intervention	A0705	Battery Problem		
					CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem		
							F08	Hospitalization or Prolonged Hospitalization	A0705	Battery Problem		
							F12	Serious Injury/ Illness/ Impairment	A0705	Battery Problem		
							F1905	Device Revision or Replacement	A0705	Battery Problem		
							F23	Unexpected Medical Intervention	A0705	Battery Problem		
					CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem		
							F08	Hospitalization or Prolonged Hospitalization	A0705	Battery Problem		
							F12	Serious Injury/ Illness/ Impairment	A0705	Battery Problem		
							F1905	Device Revision or Replacement	A0705	Battery Problem		
							F23	Unexpected Medical Intervention	A0705	Battery Problem		
					CT233	Surgical	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem		
							F08	Hospitalization or Prolonged Hospitalization	A0705	Battery Problem		
							F12	Serious Injury/ Illness/ Impairment	A0705	Battery Problem		
							F1905	Device Revision or Replacement	A0705	Battery Problem		
							F23	Unexpected Medical Intervention	A0705	Battery Problem		
					CT2406	Implantable	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem		
							F08	Hospitalization or Prolonged Hospitalization	A0705	Battery Problem		
							F12	Serious Injury/ Illness/ Impairment	A0705	Battery Problem		
							F1905	Device Revision or Replacement	A0705	Battery Problem		
							F23	Unexpected Medical Intervention	A0705	Battery Problem		
					CT2407	Active implantable	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem		
							F08	Hospitalization or Prolonged Hospitalization	A0705	Battery Problem		
							F12	Serious Injury/ Illness/ Impairment	A0705	Battery Problem		
							F1905	Device Revision or Replacement	A0705	Battery Problem		
							F23	Unexpected Medical Intervention	A0705	Battery Problem		
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem		
							F08	Hospitalization or Prolonged Hospitalization	A0705	Battery Problem		
							F12	Serious Injury/ Illness/ Impairment	A0705	Battery Problem		
							F1905	Device Revision or Replacement	A0705	Battery Problem		
							F23	Unexpected Medical Intervention	A0705	Battery Problem		
					CT2727	Active implant devices/ancillaries	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem		
							F08	Hospitalization or Prolonged Hospitalization	A0705	Battery Problem		
							F12	Serious Injury/ Illness/ Impairment	A0705	Battery Problem		
							F1905	Device Revision or Replacement	A0705	Battery Problem		
							F23	Unexpected Medical Intervention	A0705	Battery Problem		
					CT2750	Implantable cardiac electrical stimulators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem		
							F08	Hospitalization or Prolonged Hospitalization	A0705	Battery Problem		
							F12	Serious Injury/ Illness/ Impairment	A0705	Battery Problem		
							F1905	Device Revision or Replacement	A0705	Battery Problem		
							F23	Unexpected Medical Intervention	A0705	Battery Problem		
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem		
							F08	Hospitalization or Prolonged Hospitalization	A0705	Battery Problem		
							F12	Serious Injury/ Illness/ Impairment	A0705	Battery Problem		
							F1905	Device Revision or Replacement	A0705	Battery Problem		
							F23	Unexpected Medical Intervention	A0705	Battery Problem		
					CT2985	Active	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem		
							F08	Hospitalization or Prolonged Hospitalization	A0705	Battery Problem		
							F12	Serious Injury/ Illness/ Impairment	A0705	Battery Problem		
							F1905	Device Revision or Replacement	A0705	Battery Problem		
							F23	Unexpected Medical Intervention	A0705	Battery Problem		
					CT3001	Implantable metal	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem		
							F08	Hospitalization or Prolonged Hospitalization	A0705	Battery Problem		
							F12	Serious Injury/ Illness/ Impairment	A0705	Battery Problem		
							F1905	Device Revision or Replacement	A0705	Battery Problem		
							F23	Unexpected Medical Intervention	A0705	Battery Problem		
					CT3002	Implantable synthetic polymer	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem		
							F08	Hospitalization or Prolonged Hospitalization	A0705	Battery Problem		
							F12	Serious Injury/ Illness/ Impairment	A0705	Battery Problem		
							F1905	Device Revision or Replacement	A0705	Battery Problem		
							F23	Unexpected Medical Intervention	A0705	Battery Problem		
					CT3006	Implantable inorganic material	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem		
							F08	Hospitalization or Prolonged Hospitalization	A0705	Battery Problem		
							F12	Serious Injury/ Illness/ Impairment	A0705	Battery Problem		
							F1905	Device Revision or Replacement	A0705	Battery Problem		
							F23	Unexpected Medical Intervention	A0705	Battery Problem		
					CT314	Electrically powered	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem		

							F08	Hospitalization or Prolonged Hospitalization	A0705	Battery Problem		
							F12	Serious Injury/ Illness/ Impairment	A0705	Battery Problem		
							F1905	Device Revision or Replacement	A0705	Battery Problem		
							F23	Unexpected Medical Intervention	A0705	Battery Problem		
					CT321	Long-term surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem		
							F08	Hospitalization or Prolonged Hospitalization	A0705	Battery Problem		
							F12	Serious Injury/ Illness/ Impairment	A0705	Battery Problem		
							F1905	Device Revision or Replacement	A0705	Battery Problem		
							F23	Unexpected Medical Intervention	A0705	Battery Problem		
					CT449	Generators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem		
							F08	Hospitalization or Prolonged Hospitalization	A0705	Battery Problem		
							F12	Serious Injury/ Illness/ Impairment	A0705	Battery Problem		
							F1905	Device Revision or Replacement	A0705	Battery Problem		
							F23	Unexpected Medical Intervention	A0705	Battery Problem		
					CT981	Single use	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem		
							F08	Hospitalization or Prolonged Hospitalization	A0705	Battery Problem		
							F12	Serious Injury/ Illness/ Impairment	A0705	Battery Problem		
							F1905	Device Revision or Replacement	A0705	Battery Problem		
							F23	Unexpected Medical Intervention	A0705	Battery Problem		
					CT983	Surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem		
							F08	Hospitalization or Prolonged Hospitalization	A0705	Battery Problem		
							F12	Serious Injury/ Illness/ Impairment	A0705	Battery Problem		
							F1905	Device Revision or Replacement	A0705	Battery Problem		
							F23	Unexpected Medical Intervention	A0705	Battery Problem		
14597	2023	All other reportable incidents	47263	Cardiac resynchronization therapy implantable pacemaker	CT1685	Cardiac pacemakers	E2103	Device Overstimulation of Tissue	A0908	Incorrect, Inadequate or Imprecise Result or Readings	28/09/2023	28/09/2023
							F08	Hospitalization or Prolonged Hospitalization	A0908	Incorrect, Inadequate or Imprecise Result or Readings	28/09/2023	28/09/2023
							F1905	Device Revision or Replacement	A0908	Incorrect, Inadequate or Imprecise Result or Readings	28/09/2023	28/09/2023
					CT257	Implantable cardiac pacemakers	E2103	Device Overstimulation of Tissue	A0908	Incorrect, Inadequate or Imprecise Result or Readings	28/09/2023	28/09/2023
							F08	Hospitalization or Prolonged Hospitalization	A0908	Incorrect, Inadequate or Imprecise Result or Readings	28/09/2023	28/09/2023
							F1905	Device Revision or Replacement	A0908	Incorrect, Inadequate or Imprecise Result or Readings	28/09/2023	28/09/2023
					CT2866	Implantable cardiac electrical stimulators	E2103	Device Overstimulation of Tissue	A0908	Incorrect, Inadequate or Imprecise Result or Readings	28/09/2023	28/09/2023
							F08	Hospitalization or Prolonged Hospitalization	A0908	Incorrect, Inadequate or Imprecise Result or Readings	28/09/2023	28/09/2023
							F1905	Device Revision or Replacement	A0908	Incorrect, Inadequate or Imprecise Result or Readings	28/09/2023	28/09/2023
14598	2023	All other reportable incidents	63188	Left-ventricular wireless pacing system sensor/transmitter	CT145	Cardiovascular devices			A0708	Power Problem		
					CT1577	Wireless monitoring system transmitters			A0708	Power Problem		
					CT1684	Cardiac pacemakers and associated devices			A0708	Power Problem		
					CT2727	Active implant devices/ancillaries			A0708	Power Problem		
					CT2750	Implantable cardiac electrical stimulators and associated devices			A0708	Power Problem		
					CT2885	Physiological monitoring/measuring system transmitters/receivers/transceivers			A0708	Power Problem		
14599	2023	All other reportable incidents	47265	Dual-chamber implantable pacemaker, rate-responsive	CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
									A0908	Incorrect, Inadequate or Imprecise Result or Readings		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
									A0908	Incorrect, Inadequate or Imprecise Result or Readings		
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
									A0908	Incorrect, Inadequate or Imprecise Result or Readings		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
									A0908	Incorrect, Inadequate or Imprecise Result or Readings		
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
									A0908	Incorrect, Inadequate or Imprecise Result or Readings		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
									A0908	Incorrect, Inadequate or Imprecise Result or Readings		
14600	2023		47265	Dual-chamber implantable pacemaker, rate-responsive	CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	16/11/2023	16/11/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	16/11/2023	16/11/2023
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	16/11/2023	16/11/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	16/11/2023	16/11/2023
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	16/11/2023	16/11/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	16/11/2023	16/11/2023
14601	2023	All other reportable incidents	47263	Cardiac resynchronization therapy implantable pacemaker	CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
14602	2023	All other reportable incidents	47270	Cardiac resynchronization therapy implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E2104	Electric Shock	A070909	Oversensing		
									A0712	Pacing Problem		
									A071302	Inappropriate/Inadequate Shock/Stimulation		
									A090801	Signal Artifact/Noise		
							F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A070909	Oversensing		
									A0712	Pacing Problem		

								A071302	Inappropriate/Inadequate Shock/Stimulation		
								A090801	Signal Artifact/Noise		
14603	2023	All other reportable incidents	47270	Cardiac resynchronization therapy implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E060101	Asystole	A070909	Oversensing		
									A0712	Pacing Problem		
									A090801	Signal Artifact/Noise		
									A1301	Failure to Read Input Signal		
							F1001	Absence of Treatment	A070909	Oversensing		
									A0712	Pacing Problem		
									A090801	Signal Artifact/Noise		
									A1301	Failure to Read Input Signal		
14604	2023	All other reportable incidents	47265	Dual-chamber implantable pacemaker, rate-responsive	CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A070909	Oversensing	11/12/2023	11/12/2023
									A0908	Incorrect, Inadequate or Imprecise Result or Readings	11/12/2023	11/12/2023
									A090801	Signal Artifact/Noise	11/12/2023	11/12/2023
									A0909	Unexpected Therapeutic Results	11/12/2023	11/12/2023
									A23	Use of Device Problem	11/12/2023	11/12/2023
							F08	Hospitalization or Prolonged Hospitalization	A070909	Oversensing	11/12/2023	11/12/2023
									A0908	Incorrect, Inadequate or Imprecise Result or Readings	11/12/2023	11/12/2023
									A090801	Signal Artifact/Noise	11/12/2023	11/12/2023
									A0909	Unexpected Therapeutic Results	11/12/2023	11/12/2023
									A23	Use of Device Problem	11/12/2023	11/12/2023
							F1905	Device Revision or Replacement	A070909	Oversensing	11/12/2023	11/12/2023
									A0908	Incorrect, Inadequate or Imprecise Result or Readings	11/12/2023	11/12/2023
									A090801	Signal Artifact/Noise	11/12/2023	11/12/2023
									A0909	Unexpected Therapeutic Results	11/12/2023	11/12/2023
									A23	Use of Device Problem	11/12/2023	11/12/2023
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A070909	Oversensing	11/12/2023	11/12/2023
									A0908	Incorrect, Inadequate or Imprecise Result or Readings	11/12/2023	11/12/2023
									A090801	Signal Artifact/Noise	11/12/2023	11/12/2023
									A0909	Unexpected Therapeutic Results	11/12/2023	11/12/2023
									A23	Use of Device Problem	11/12/2023	11/12/2023
							F08	Hospitalization or Prolonged Hospitalization	A070909	Oversensing	11/12/2023	11/12/2023
									A0908	Incorrect, Inadequate or Imprecise Result or Readings	11/12/2023	11/12/2023
									A090801	Signal Artifact/Noise	11/12/2023	11/12/2023
									A0909	Unexpected Therapeutic Results	11/12/2023	11/12/2023
									A23	Use of Device Problem	11/12/2023	11/12/2023
							F1905	Device Revision or Replacement	A070909	Oversensing	11/12/2023	11/12/2023
									A0908	Incorrect, Inadequate or Imprecise Result or Readings	11/12/2023	11/12/2023
									A090801	Signal Artifact/Noise	11/12/2023	11/12/2023
									A0909	Unexpected Therapeutic Results	11/12/2023	11/12/2023
									A23	Use of Device Problem	11/12/2023	11/12/2023
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A070909	Oversensing	11/12/2023	11/12/2023
									A0908	Incorrect, Inadequate or Imprecise Result or Readings	11/12/2023	11/12/2023
									A090801	Signal Artifact/Noise	11/12/2023	11/12/2023
									A0909	Unexpected Therapeutic Results	11/12/2023	11/12/2023
									A23	Use of Device Problem	11/12/2023	11/12/2023
							F08	Hospitalization or Prolonged Hospitalization	A070909	Oversensing	11/12/2023	11/12/2023
									A0908	Incorrect, Inadequate or Imprecise Result or Readings	11/12/2023	11/12/2023
									A090801	Signal Artifact/Noise	11/12/2023	11/12/2023
									A0909	Unexpected Therapeutic Results	11/12/2023	11/12/2023
									A23	Use of Device Problem	11/12/2023	11/12/2023
							F1905	Device Revision or Replacement	A070909	Oversensing	11/12/2023	11/12/2023
									A0908	Incorrect, Inadequate or Imprecise Result or Readings	11/12/2023	11/12/2023
									A090801	Signal Artifact/Noise	11/12/2023	11/12/2023
									A0909	Unexpected Therapeutic Results	11/12/2023	11/12/2023
									A23	Use of Device Problem	11/12/2023	11/12/2023
14605	2023	All other reportable incidents	60789	Intracardiac pacemaker	CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A051201	Device Dislodged or Dislocated		
									A070101	Failure to Capture		
									A0722	Impedance Problem		
							F26	No Health Consequences or Impact	A051201	Device Dislodged or Dislocated		
									A070101	Failure to Capture		
									A0722	Impedance Problem		
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A051201	Device Dislodged or Dislocated		
									A070101	Failure to Capture		
									A0722	Impedance Problem		
							F26	No Health Consequences or Impact	A051201	Device Dislodged or Dislocated		
									A070101	Failure to Capture		
									A0722	Impedance Problem		
					CT2818	Catheterization/Cannulation/Fluid path devices	E2403	No Clinical Signs, Symptoms or Conditions	A051201	Device Dislodged or Dislocated		
									A070101	Failure to Capture		
									A0722	Impedance Problem		
							F26	No Health Consequences or Impact	A051201	Device Dislodged or Dislocated		
									A070101	Failure to Capture		
									A0722	Impedance Problem		
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A051201	Device Dislodged or Dislocated		
									A070101	Failure to Capture		
									A0722	Impedance Problem		
							F26	No Health Consequences or Impact	A051201	Device Dislodged or Dislocated		
									A070101	Failure to Capture		
									A0722	Impedance Problem		
14606	2023	All other reportable incidents	17882	Physiologic-monitoring defibrillation system	CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A0719	Unexpected Shutdown		
							F26	No Health Consequences or Impact	A0719	Unexpected Shutdown		

14607	2023	All other reportable incidents	47265	Dual-chamber implantable pacemaker, rate-responsive	CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A0723	Unexpected Mode Switch	15/01/2024	15/01/2024
							F1901	Additional Surgery	A0723	Unexpected Mode Switch	15/01/2024	15/01/2024
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A0723	Unexpected Mode Switch	15/01/2024	15/01/2024
							F1901	Additional Surgery	A0723	Unexpected Mode Switch	15/01/2024	15/01/2024
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A0723	Unexpected Mode Switch	15/01/2024	15/01/2024
							F1901	Additional Surgery	A0723	Unexpected Mode Switch	15/01/2024	15/01/2024
14608	2023		47265	Dual-chamber implantable pacemaker, rate-responsive	CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	13/12/2023	13/12/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	13/12/2023	13/12/2023
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	13/12/2023	13/12/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	13/12/2023	13/12/2023
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	13/12/2023	13/12/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	13/12/2023	13/12/2023
14609	2023		47265	Dual-chamber implantable pacemaker, rate-responsive	CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	18/12/2023	18/12/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	18/12/2023	18/12/2023
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	18/12/2023	18/12/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	18/12/2023	18/12/2023
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	18/12/2023	18/12/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	18/12/2023	18/12/2023
14610	2023	All other reportable incidents	35852	Single-chamber implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A070909	Oversensing		
									A070910	Undersensing		
									A072202	Low Impedance		
									A090801	Signal Artifact/Noise		
									A1301	Failure to Read Input Signal		
							F26	No Health Consequences or Impact	A070909	Oversensing		
									A070910	Undersensing		
									A072202	Low Impedance		
									A090801	Signal Artifact/Noise		
									A1301	Failure to Read Input Signal		
14611	2023	All other reportable incidents	47265	Dual-chamber implantable pacemaker, rate-responsive	CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A0908	Incorrect, Inadequate or Imprecise Result or Readings	15/12/2023	15/12/2023
							F08	Hospitalization or Prolonged Hospitalization	A0908	Incorrect, Inadequate or Imprecise Result or Readings	15/12/2023	15/12/2023
							F1905	Device Revision or Replacement	A0908	Incorrect, Inadequate or Imprecise Result or Readings	15/12/2023	15/12/2023
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A0908	Incorrect, Inadequate or Imprecise Result or Readings	15/12/2023	15/12/2023
							F08	Hospitalization or Prolonged Hospitalization	A0908	Incorrect, Inadequate or Imprecise Result or Readings	15/12/2023	15/12/2023
							F1905	Device Revision or Replacement	A0908	Incorrect, Inadequate or Imprecise Result or Readings	15/12/2023	15/12/2023
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A0908	Incorrect, Inadequate or Imprecise Result or Readings	15/12/2023	15/12/2023
							F08	Hospitalization or Prolonged Hospitalization	A0908	Incorrect, Inadequate or Imprecise Result or Readings	15/12/2023	15/12/2023
							F1905	Device Revision or Replacement	A0908	Incorrect, Inadequate or Imprecise Result or Readings	15/12/2023	15/12/2023
14612	2023	All other reportable incidents	47205	Cardiac pulse generator programmer	CT1684	Cardiac pacemakers and associated devices	E0601	Arrhythmia	A110303	Program or Algorithm Execution Failure		
							F26	No Health Consequences or Impact	A110303	Program or Algorithm Execution Failure		
14613	2023		47265	Dual-chamber implantable pacemaker, rate-responsive	CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	09/12/2023	09/12/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	09/12/2023	09/12/2023
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	09/12/2023	09/12/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	09/12/2023	09/12/2023
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	09/12/2023	09/12/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	09/12/2023	09/12/2023
14614	2023		47265	Dual-chamber implantable pacemaker, rate-responsive	CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	09/12/2023	09/12/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	09/12/2023	09/12/2023
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	09/12/2023	09/12/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	09/12/2023	09/12/2023
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	09/12/2023	09/12/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	09/12/2023	09/12/2023
14615	2023		47265	Dual-chamber implantable pacemaker, rate-responsive	CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	09/12/2023	09/12/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	09/12/2023	09/12/2023
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	09/12/2023	09/12/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	09/12/2023	09/12/2023
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	09/12/2023	09/12/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	09/12/2023	09/12/2023
14616	2023		47265	Dual-chamber implantable pacemaker, rate-responsive	CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	09/12/2023	09/12/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	09/12/2023	09/12/2023
							F1903	Device Explantation	A24	No Apparent Device Problem	09/12/2023	09/12/2023
							F1905	Device Revision or Replacement	A24	No Apparent Device Problem	09/12/2023	09/12/2023
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	09/12/2023	09/12/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	09/12/2023	09/12/2023
							F1903	Device Explantation	A24	No Apparent Device Problem	09/12/2023	09/12/2023
							F1905	Device Revision or Replacement	A24	No Apparent Device Problem	09/12/2023	09/12/2023
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	09/12/2023	09/12/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	09/12/2023	09/12/2023
							F1903	Device Explantation	A24	No Apparent Device Problem	09/12/2023	09/12/2023
							F1905	Device Revision or Replacement	A24	No Apparent Device Problem	09/12/2023	09/12/2023
14617	2023	All other reportable incidents	47265	Dual-chamber implantable pacemaker, rate-responsive	CT1685	Cardiac pacemakers	E2103	Device Overstimulation of Tissue	A070504	Premature Discharge of Battery		
									A071209	Extracardiac Stimulation		
							F12	Serious Injury/ Illness/ Impairment	A070504	Premature Discharge of Battery		
									A071209	Extracardiac Stimulation		
							F23	Unexpected Medical Intervention	A070504	Premature Discharge of Battery		
									A071209	Extracardiac Stimulation		
					CT257	Implantable cardiac pacemakers	E2103	Device Overstimulation of Tissue	A070504	Premature Discharge of Battery		
									A071209	Extracardiac Stimulation		
							F12	Serious Injury/ Illness/ Impairment	A070504	Premature Discharge of Battery		
									A071209	Extracardiac Stimulation		
							F23	Unexpected Medical Intervention	A070504	Premature Discharge of Battery		
									A071209	Extracardiac Stimulation		

					CT2866	Implantable cardiac electrical stimulators	E2103	Device Overstimulation of Tissue	A070504	Premature Discharge of Battery		
									A071209	Extracardiac Stimulation		
							F12	Serious Injury/ Illness/ Impairment	A070504	Premature Discharge of Battery		
									A071209	Extracardiac Stimulation		
							F23	Unexpected Medical Intervention	A070504	Premature Discharge of Battery		
									A071209	Extracardiac Stimulation		
14618	2023	All other reportable incidents	47270	Cardiac resynchronization therapy implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A072201	High Impedance	07/12/2023	07/12/2023
							F08	Hospitalization or Prolonged Hospitalization	A072201	High Impedance	07/12/2023	07/12/2023
							F1905	Device Revision or Replacement	A072201	High Impedance	07/12/2023	07/12/2023
14619	2023		47265	Dual-chamber implantable pacemaker, rate-responsive	CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem		
							F1901	Additional Surgery	A24	No Apparent Device Problem		
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem		
							F1901	Additional Surgery	A24	No Apparent Device Problem		
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem		
							F1901	Additional Surgery	A24	No Apparent Device Problem		
14620	2023	All other reportable incidents	17882	Physiologic-monitoring defibrillation system	CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A0719	Unexpected Shutdown		
							F27	Problem Identified During Non-Clinical Procedure	A0719	Unexpected Shutdown		
14621	2023	All other reportable incidents	17882	Physiologic-monitoring defibrillation system	CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070803	Failure to Power Up		
							F27	Problem Identified During Non-Clinical Procedure	A070803	Failure to Power Up		
14622	2023	All other reportable incidents	17882	Physiologic-monitoring defibrillation system	CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A071301	Failure to Deliver Shock/Stimulation		
									A150105	Key or Button Unresponsive/not Working		
							F27	Problem Identified During Non-Clinical Procedure	A071301	Failure to Deliver Shock/Stimulation		
									A150105	Key or Button Unresponsive/not Working		
14623	2023	All other reportable incidents	37265	Dual-chamber implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E2104	Electric Shock	A071302	Inappropriate/Inadequate Shock/Stimulation		
							F26	No Health Consequences or Impact	A071302	Inappropriate/Inadequate Shock/Stimulation		
14624	2023	All other reportable incidents	36241	Pacemaker/defibrillator lead	CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070910	Undersensing		
									A090801	Signal Artifact/Noise		
							F26	No Health Consequences or Impact	A070910	Undersensing		
									A090801	Signal Artifact/Noise		
14625	2023		47267	Single-chamber implantable pacemaker, rate-responsive	CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	30/11/2023	30/11/2023
							F1901	Additional Surgery	A070504	Premature Discharge of Battery	30/11/2023	30/11/2023
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	30/11/2023	30/11/2023
							F1901	Additional Surgery	A070504	Premature Discharge of Battery	30/11/2023	30/11/2023
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	30/11/2023	30/11/2023
							F1901	Additional Surgery	A070504	Premature Discharge of Battery	30/11/2023	30/11/2023
14626	2023	All other reportable incidents	17882	Physiologic-monitoring defibrillation system	CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A0719	Unexpected Shutdown		
							F27	Problem Identified During Non-Clinical Procedure	A0719	Unexpected Shutdown		
14627	2023	All other reportable incidents	17882	Physiologic-monitoring defibrillation system	CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A071204	No Pacing		
							F27	Problem Identified During Non-Clinical Procedure	A071204	No Pacing		
14628	2023	All other reportable incidents	47205	Cardiac pulse generator programmer	CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A13	Communication or Transmission Problem		
							F23	Unexpected Medical Intervention	A13	Communication or Transmission Problem		
14629	2023	All other reportable incidents	35223	Endocardial/interventricular septal pacing lead	CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A051201	Device Dislodged or Dislocated	03/01/2024	03/01/2024
									A0701	Capturing Problem	03/01/2024	03/01/2024
									A070101	Failure to Capture	03/01/2024	03/01/2024
									A070102	High Capture Threshold	03/01/2024	03/01/2024
									A070901	Decreased Sensitivity	03/01/2024	03/01/2024
									A072201	High Impedance	03/01/2024	03/01/2024
							F08	Hospitalization or Prolonged Hospitalization	A051201	Device Dislodged or Dislocated	03/01/2024	03/01/2024
									A0701	Capturing Problem	03/01/2024	03/01/2024
									A070101	Failure to Capture	03/01/2024	03/01/2024
									A070102	High Capture Threshold	03/01/2024	03/01/2024
									A070901	Decreased Sensitivity	03/01/2024	03/01/2024
									A072201	High Impedance	03/01/2024	03/01/2024
							F12	Serious Injury/ Illness/ Impairment	A051201	Device Dislodged or Dislocated	03/01/2024	03/01/2024
									A0701	Capturing Problem	03/01/2024	03/01/2024
									A070101	Failure to Capture	03/01/2024	03/01/2024
									A070102	High Capture Threshold	03/01/2024	03/01/2024
									A070901	Decreased Sensitivity	03/01/2024	03/01/2024
									A072201	High Impedance	03/01/2024	03/01/2024
							F1903	Device Explantation	A051201	Device Dislodged or Dislocated	03/01/2024	03/01/2024
									A0701	Capturing Problem	03/01/2024	03/01/2024
									A070101	Failure to Capture	03/01/2024	03/01/2024
									A070102	High Capture Threshold	03/01/2024	03/01/2024
									A070901	Decreased Sensitivity	03/01/2024	03/01/2024
									A072201	High Impedance	03/01/2024	03/01/2024
							F1905	Device Revision or Replacement	A051201	Device Dislodged or Dislocated	03/01/2024	03/01/2024
									A0701	Capturing Problem	03/01/2024	03/01/2024
									A070101	Failure to Capture	03/01/2024	03/01/2024
									A070102	High Capture Threshold	03/01/2024	03/01/2024
									A070901	Decreased Sensitivity	03/01/2024	03/01/2024
									A072201	High Impedance	03/01/2024	03/01/2024
							F23	Unexpected Medical Intervention	A051201	Device Dislodged or Dislocated	03/01/2024	03/01/2024
									A0701	Capturing Problem	03/01/2024	03/01/2024
									A070101	Failure to Capture	03/01/2024	03/01/2024
									A070102	High Capture Threshold	03/01/2024	03/01/2024
									A070901	Decreased Sensitivity	03/01/2024	03/01/2024
									A072201	High Impedance	03/01/2024	03/01/2024
14630	2023	All other reportable incidents	47263	Cardiac resynchronization therapy implantable pacemaker	CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem		
									A072202	Low Impedance		
							F26	No Health Consequences or Impact	A0705	Battery Problem		
									A072202	Low Impedance		
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem		
									A072202	Low Impedance		
							F26	No Health Consequences or Impact	A0705	Battery Problem		
									A072202	Low Impedance		
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem		
									A072202	Low Impedance		
							F26	No Health Consequences or Impact	A0705	Battery Problem		

									A072202	Low Impedance		
14631	2023	All other reportable incidents	47265	Dual-chamber implantable pacemaker, rate-responsive	CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
14632	2023		47205	Cardiac pulse generator programmer	CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem		
									A1601	Device Alarm System		
									A160601	Premature Elective Replacement Indicator		
							F1903	Device Explantation	A0705	Battery Problem		
									A1601	Device Alarm System		
									A160601	Premature Elective Replacement Indicator		
14633	2023		47265	Dual-chamber implantable pacemaker, rate-responsive	CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	17/11/2023	17/11/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	17/11/2023	17/11/2023
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	17/11/2023	17/11/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	17/11/2023	17/11/2023
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	17/11/2023	17/11/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	17/11/2023	17/11/2023
14634	2023	All other reportable incidents	37265	Dual-chamber implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem	05/06/2023	05/06/2023
									A070504	Premature Discharge of Battery	05/06/2023	05/06/2023
							F08	Hospitalization or Prolonged Hospitalization	A0705	Battery Problem	05/06/2023	05/06/2023
									A070504	Premature Discharge of Battery	05/06/2023	05/06/2023
							F1905	Device Revision or Replacement	A0705	Battery Problem	05/06/2023	05/06/2023
									A070504	Premature Discharge of Battery	05/06/2023	05/06/2023
14635	2023	All other reportable incidents	35852	Single-chamber implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	13/12/2023	13/12/2023
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	13/12/2023	13/12/2023
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	13/12/2023	13/12/2023
14636	2023	All other reportable incidents	47265	Dual-chamber implantable pacemaker, rate-responsive	CT141	Electrodes/Leads and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	16/09/2022	16/09/2022
							F08	Hospitalization or Prolonged Hospitalization	A24	No Apparent Device Problem	16/09/2022	16/09/2022
							F1905	Device Revision or Replacement	A24	No Apparent Device Problem	16/09/2022	16/09/2022
					CT145	Cardiovascular devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	16/09/2022	16/09/2022
							F08	Hospitalization or Prolonged Hospitalization	A24	No Apparent Device Problem	16/09/2022	16/09/2022
							F1905	Device Revision or Replacement	A24	No Apparent Device Problem	16/09/2022	16/09/2022
					CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	16/09/2022	16/09/2022
							F08	Hospitalization or Prolonged Hospitalization	A24	No Apparent Device Problem	16/09/2022	16/09/2022
							F1905	Device Revision or Replacement	A24	No Apparent Device Problem	16/09/2022	16/09/2022
					CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	16/09/2022	16/09/2022
							F08	Hospitalization or Prolonged Hospitalization	A24	No Apparent Device Problem	16/09/2022	16/09/2022
							F1905	Device Revision or Replacement	A24	No Apparent Device Problem	16/09/2022	16/09/2022
					CT233	Surgical	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	16/09/2022	16/09/2022
							F08	Hospitalization or Prolonged Hospitalization	A24	No Apparent Device Problem	16/09/2022	16/09/2022
							F1905	Device Revision or Replacement	A24	No Apparent Device Problem	16/09/2022	16/09/2022
					CT2406	Implantable	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	16/09/2022	16/09/2022
							F08	Hospitalization or Prolonged Hospitalization	A24	No Apparent Device Problem	16/09/2022	16/09/2022
							F1905	Device Revision or Replacement	A24	No Apparent Device Problem	16/09/2022	16/09/2022
					CT2407	Active implantable	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	16/09/2022	16/09/2022
							F08	Hospitalization or Prolonged Hospitalization	A24	No Apparent Device Problem	16/09/2022	16/09/2022
							F1905	Device Revision or Replacement	A24	No Apparent Device Problem	16/09/2022	16/09/2022
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	16/09/2022	16/09/2022
							F08	Hospitalization or Prolonged Hospitalization	A24	No Apparent Device Problem	16/09/2022	16/09/2022
							F1905	Device Revision or Replacement	A24	No Apparent Device Problem	16/09/2022	16/09/2022
					CT2727	Active implant devices/ancillaries	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	16/09/2022	16/09/2022
							F08	Hospitalization or Prolonged Hospitalization	A24	No Apparent Device Problem	16/09/2022	16/09/2022
							F1905	Device Revision or Replacement	A24	No Apparent Device Problem	16/09/2022	16/09/2022
					CT2750	Implantable cardiac electrical stimulators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	16/09/2022	16/09/2022
							F08	Hospitalization or Prolonged Hospitalization	A24	No Apparent Device Problem	16/09/2022	16/09/2022
							F1905	Device Revision or Replacement	A24	No Apparent Device Problem	16/09/2022	16/09/2022
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	16/09/2022	16/09/2022
							F08	Hospitalization or Prolonged Hospitalization	A24	No Apparent Device Problem	16/09/2022	16/09/2022
							F1905	Device Revision or Replacement	A24	No Apparent Device Problem	16/09/2022	16/09/2022
					CT2985	Active	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	16/09/2022	16/09/2022
							F08	Hospitalization or Prolonged Hospitalization	A24	No Apparent Device Problem	16/09/2022	16/09/2022
							F1905	Device Revision or Replacement	A24	No Apparent Device Problem	16/09/2022	16/09/2022
					CT3001	Implantable metal	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	16/09/2022	16/09/2022
							F08	Hospitalization or Prolonged Hospitalization	A24	No Apparent Device Problem	16/09/2022	16/09/2022
							F1905	Device Revision or Replacement	A24	No Apparent Device Problem	16/09/2022	16/09/2022
					CT3002	Implantable synthetic polymer	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	16/09/2022	16/09/2022
							F08	Hospitalization or Prolonged Hospitalization	A24	No Apparent Device Problem	16/09/2022	16/09/2022
							F1905	Device Revision or Replacement	A24	No Apparent Device Problem	16/09/2022	16/09/2022
					CT3006	Implantable inorganic material	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	16/09/2022	16/09/2022
							F08	Hospitalization or Prolonged Hospitalization	A24	No Apparent Device Problem	16/09/2022	16/09/2022
							F1905	Device Revision or Replacement	A24	No Apparent Device Problem	16/09/2022	16/09/2022
					CT314	Electrically powered	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	16/09/2022	16/09/2022
							F08	Hospitalization or Prolonged Hospitalization	A24	No Apparent Device Problem	16/09/2022	16/09/2022
							F1905	Device Revision or Replacement	A24	No Apparent Device Problem	16/09/2022	16/09/2022
					CT321	Long-term surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	16/09/2022	16/09/2022
							F08	Hospitalization or Prolonged Hospitalization	A24	No Apparent Device Problem	16/09/2022	16/09/2022
							F1905	Device Revision or Replacement	A24	No Apparent Device Problem	16/09/2022	16/09/2022
					CT449	Generators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	16/09/2022	16/09/2022
							F08	Hospitalization or Prolonged Hospitalization	A24	No Apparent Device Problem	16/09/2022	16/09/2022
							F1905	Device Revision or Replacement	A24	No Apparent Device Problem	16/09/2022	16/09/2022

					CT981	Single use	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	16/09/2022	16/09/2022
							F08	Hospitalization or Prolonged Hospitalization	A24	No Apparent Device Problem	16/09/2022	16/09/2022
							F1905	Device Revision or Replacement	A24	No Apparent Device Problem	16/09/2022	16/09/2022
					CT983	Surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	16/09/2022	16/09/2022
							F08	Hospitalization or Prolonged Hospitalization	A24	No Apparent Device Problem	16/09/2022	16/09/2022
							F1905	Device Revision or Replacement	A24	No Apparent Device Problem	16/09/2022	16/09/2022
14637	2023	All other reportable incidents	17882	Physiologic-monitoring defibrillation system	CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A12	Connection Problem		
							F27	Problem Identified During Non-Clinical Procedure	A12	Connection Problem		
14638	2023	All other reportable incidents	17882	Physiologic-monitoring defibrillation system	CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070803	Failure to Power Up		
									A0719	Unexpected Shutdown		
							F26	No Health Consequences or Impact	A070803	Failure to Power Up		
									A0719	Unexpected Shutdown		
14639	2023	All other reportable incidents	17882	Physiologic-monitoring defibrillation system	CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A0713	Defibrillation/Stimulation Problem		
							F26	No Health Consequences or Impact	A0713	Defibrillation/Stimulation Problem		
14640	2023	All other reportable incidents	47265	Dual-chamber implantable pacemaker, rate-responsive	CT1685	Cardiac pacemakers	E0112	Dizziness	A070101	Failure to Capture	21/12/2023	21/12/2023
									A0723	Unexpected Mode Switch	21/12/2023	21/12/2023
									A090407	Therapy Delivered to Incorrect Body Area	21/12/2023	21/12/2023
							E2311	Discomfort	A070101	Failure to Capture	21/12/2023	21/12/2023
									A0723	Unexpected Mode Switch	21/12/2023	21/12/2023
									A090407	Therapy Delivered to Incorrect Body Area	21/12/2023	21/12/2023
							F1901	Additional Surgery	A070101	Failure to Capture	21/12/2023	21/12/2023
									A0723	Unexpected Mode Switch	21/12/2023	21/12/2023
									A090407	Therapy Delivered to Incorrect Body Area	21/12/2023	21/12/2023
					CT257	Implantable cardiac pacemakers	E0112	Dizziness	A070101	Failure to Capture	21/12/2023	21/12/2023
									A0723	Unexpected Mode Switch	21/12/2023	21/12/2023
									A090407	Therapy Delivered to Incorrect Body Area	21/12/2023	21/12/2023
							E2311	Discomfort	A070101	Failure to Capture	21/12/2023	21/12/2023
									A0723	Unexpected Mode Switch	21/12/2023	21/12/2023
									A090407	Therapy Delivered to Incorrect Body Area	21/12/2023	21/12/2023
							F1901	Additional Surgery	A070101	Failure to Capture	21/12/2023	21/12/2023
									A0723	Unexpected Mode Switch	21/12/2023	21/12/2023
									A090407	Therapy Delivered to Incorrect Body Area	21/12/2023	21/12/2023
					CT2866	Implantable cardiac electrical stimulators	E0112	Dizziness	A070101	Failure to Capture	21/12/2023	21/12/2023
									A0723	Unexpected Mode Switch	21/12/2023	21/12/2023
									A090407	Therapy Delivered to Incorrect Body Area	21/12/2023	21/12/2023
							E2311	Discomfort	A070101	Failure to Capture	21/12/2023	21/12/2023
									A0723	Unexpected Mode Switch	21/12/2023	21/12/2023
									A090407	Therapy Delivered to Incorrect Body Area	21/12/2023	21/12/2023
							F1901	Additional Surgery	A070101	Failure to Capture	21/12/2023	21/12/2023
									A0723	Unexpected Mode Switch	21/12/2023	21/12/2023
									A090407	Therapy Delivered to Incorrect Body Area	21/12/2023	21/12/2023
14641	2023	All other reportable incidents	17882	Physiologic-monitoring defibrillation system	CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A0908	Incorrect, Inadequate or Imprecise Result or Readings		
							F26	No Health Consequences or Impact	A0908	Incorrect, Inadequate or Imprecise Result or Readings		
14642	2023	All other reportable incidents	60789	Intracardiac pacemaker	CT1685	Cardiac pacemakers	E0204	Lethargy	A070101	Failure to Capture		
									A070102	High Capture Threshold		
									A070909	Oversensing		
							E0601	Arrhythmia	A070101	Failure to Capture		
									A070102	High Capture Threshold		
									A070909	Oversensing		
							E060104	Bradycardia	A070101	Failure to Capture		
									A070102	High Capture Threshold		
									A070909	Oversensing		
							F08	Hospitalization or Prolonged Hospitalization	A070101	Failure to Capture		
									A070102	High Capture Threshold		
									A070909	Oversensing		
							F12	Serious Injury/ Illness/ Impairment	A070101	Failure to Capture		
									A070102	High Capture Threshold		
									A070909	Oversensing		
					CT257	Implantable cardiac pacemakers	E0204	Lethargy	A070101	Failure to Capture		
									A070102	High Capture Threshold		
									A070909	Oversensing		
							E0601	Arrhythmia	A070101	Failure to Capture		
									A070102	High Capture Threshold		
									A070909	Oversensing		
							E060104	Bradycardia	A070101	Failure to Capture		
									A070102	High Capture Threshold		
									A070909	Oversensing		
							F08	Hospitalization or Prolonged Hospitalization	A070101	Failure to Capture		
									A070102	High Capture Threshold		
									A070909	Oversensing		
							F12	Serious Injury/ Illness/ Impairment	A070101	Failure to Capture		
									A070102	High Capture Threshold		
									A070909	Oversensing		
					CT2818	Catheterization/Cannulation/Fluid path devices	E0204	Lethargy	A070101	Failure to Capture		
									A070102	High Capture Threshold		
									A070909	Oversensing		
							E0601	Arrhythmia	A070101	Failure to Capture		
									A070102	High Capture Threshold		
									A070909	Oversensing		
							E060104	Bradycardia	A070101	Failure to Capture		
									A070102	High Capture Threshold		
									A070909	Oversensing		
							F08	Hospitalization or Prolonged Hospitalization	A070101	Failure to Capture		
									A070102	High Capture Threshold		
									A070909	Oversensing		
							F12	Serious Injury/ Illness/ Impairment	A070101	Failure to Capture		

									A070102	High Capture Threshold		
									A070909	Oversensing		
					CT2866	Implantable cardiac electrical stimulators	E0204	Lethargy	A070101	Failure to Capture		
									A070102	High Capture Threshold		
									A070909	Oversensing		
							E0601	Arrhythmia	A070101	Failure to Capture		
									A070102	High Capture Threshold		
									A070909	Oversensing		
							E060104	Bradycardia	A070101	Failure to Capture		
									A070102	High Capture Threshold		
									A070909	Oversensing		
							F08	Hospitalization or Prolonged Hospitalization	A070101	Failure to Capture		
									A070102	High Capture Threshold		
									A070909	Oversensing		
							F12	Serious Injury/ Illness/ Impairment	A070101	Failure to Capture		
									A070102	High Capture Threshold		
									A070909	Oversensing		
14643	2023	All other reportable incidents	60789	Intracardiac pacemaker	CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT2818	Catheterization/Cannulation/Fluid path devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
14644	2023	All other reportable incidents	37265	Dual-chamber implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E060106	Heart Block	A0712	Pacing Problem		
							F12	Serious Injury/ Illness/ Impairment	A0712	Pacing Problem		
							F23	Unexpected Medical Intervention	A0712	Pacing Problem		
14645	2023		47263	Cardiac resynchronization therapy implantable pacemaker	CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	13/06/2023	13/06/2023
							F1901	Additional Surgery	A070504	Premature Discharge of Battery	13/06/2023	13/06/2023
					CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	13/06/2023	13/06/2023
							F1901	Additional Surgery	A070504	Premature Discharge of Battery	13/06/2023	13/06/2023
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	13/06/2023	13/06/2023
							F1901	Additional Surgery	A070504	Premature Discharge of Battery	13/06/2023	13/06/2023
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	13/06/2023	13/06/2023
							F1901	Additional Surgery	A070504	Premature Discharge of Battery	13/06/2023	13/06/2023
14646	2023	All other reportable incidents	37265	Dual-chamber implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A0709	Device Sensing Problem		
									A070909	Oversensing		
									A090801	Signal Artifact/Noise		
							F26	No Health Consequences or Impact	A0709	Device Sensing Problem		
									A070909	Oversensing		
									A090801	Signal Artifact/Noise		
14647	2023	All other reportable incidents	17882	Physiologic-monitoring defibrillation system	CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A090811	Unable to Obtain Readings		
							F26	No Health Consequences or Impact	A090811	Unable to Obtain Readings		
14648	2023	All other reportable incidents	35852	Single-chamber implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	12/12/2023	12/12/2023
									A0908	Incorrect, Inadequate or Imprecise Result or Readings	12/12/2023	12/12/2023
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	12/12/2023	12/12/2023
									A0908	Incorrect, Inadequate or Imprecise Result or Readings	12/12/2023	12/12/2023
							F1903	Device Explantation	A070504	Premature Discharge of Battery	12/12/2023	12/12/2023
									A0908	Incorrect, Inadequate or Imprecise Result or Readings	12/12/2023	12/12/2023
14649	2023	All other reportable incidents	47265	Dual-chamber implantable pacemaker, rate-responsive	CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
14650	2023	All other reportable incidents	47265	Dual-chamber implantable pacemaker, rate-responsive	CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A0712	Pacing Problem		
							F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A0712	Pacing Problem		
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A0712	Pacing Problem		
							F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A0712	Pacing Problem		
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A0712	Pacing Problem		
							F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A0712	Pacing Problem		
14651	2023	All other reportable incidents	47265	Dual-chamber implantable pacemaker, rate-responsive	CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem	19/12/2023	19/12/2023
									A070504	Premature Discharge of Battery	19/12/2023	19/12/2023
							F08	Hospitalization or Prolonged Hospitalization	A0705	Battery Problem	19/12/2023	19/12/2023
									A070504	Premature Discharge of Battery	19/12/2023	19/12/2023
							F1905	Device Revision or Replacement	A0705	Battery Problem	19/12/2023	19/12/2023
									A070504	Premature Discharge of Battery	19/12/2023	19/12/2023
							F23	Unexpected Medical Intervention	A0705	Battery Problem	19/12/2023	19/12/2023
									A070504	Premature Discharge of Battery	19/12/2023	19/12/2023
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem	19/12/2023	19/12/2023
									A070504	Premature Discharge of Battery	19/12/2023	19/12/2023
							F08	Hospitalization or Prolonged Hospitalization	A0705	Battery Problem	19/12/2023	19/12/2023
									A070504	Premature Discharge of Battery	19/12/2023	19/12/2023
							F1905	Device Revision or Replacement	A0705	Battery Problem	19/12/2023	19/12/2023
									A070504	Premature Discharge of Battery	19/12/2023	19/12/2023
							F23	Unexpected Medical Intervention	A0705	Battery Problem	19/12/2023	19/12/2023
									A070504	Premature Discharge of Battery	19/12/2023	19/12/2023
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem	19/12/2023	19/12/2023

									A070504	Premature Discharge of Battery	19/12/2023	19/12/2023
							F08	Hospitalization or Prolonged Hospitalization	A0705	Battery Problem	19/12/2023	19/12/2023
									A070504	Premature Discharge of Battery	19/12/2023	19/12/2023
							F1905	Device Revision or Replacement	A0705	Battery Problem	19/12/2023	19/12/2023
									A070504	Premature Discharge of Battery	19/12/2023	19/12/2023
							F23	Unexpected Medical Intervention	A0705	Battery Problem	19/12/2023	19/12/2023
									A070504	Premature Discharge of Battery	19/12/2023	19/12/2023
14652	2023	All other reportable incidents	47265	Dual-chamber implantable pacemaker, rate-responsive	CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
14653	2023	All other reportable incidents			CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	01/12/2023	01/12/2023
									A0908	Incorrect, Inadequate or Imprecise Result or Readings	01/12/2023	01/12/2023
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	01/12/2023	01/12/2023
									A0908	Incorrect, Inadequate or Imprecise Result or Readings	01/12/2023	01/12/2023
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	01/12/2023	01/12/2023
									A0908	Incorrect, Inadequate or Imprecise Result or Readings	01/12/2023	01/12/2023
14654	2023	All other reportable incidents	47270	Cardiac resynchronization therapy implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E060101	Asystole	A070909	Oversensing		
									A0712	Pacing Problem		
									A090801	Signal Artifact/Noise		
							F1001	Absence of Treatment	A070909	Oversensing		
									A0712	Pacing Problem		
									A090801	Signal Artifact/Noise		
14655	2023	All other reportable incidents	37265	Dual-chamber implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E2104	Electric Shock	A070910	Undersensing		
									A0712	Pacing Problem		
									A071302	Inappropriate/Inadequate Shock/Stimulation		
							F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A070910	Undersensing		
									A0712	Pacing Problem		
									A071302	Inappropriate/Inadequate Shock/Stimulation		
14656	2023	All other reportable incidents	37265	Dual-chamber implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	01/08/2024	01/08/2024
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	01/08/2024	01/08/2024
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	01/08/2024	01/08/2024
14657	2023	All other reportable incidents	37265	Dual-chamber implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	21/12/2023	21/12/2023
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	21/12/2023	21/12/2023
							F12	Serious Injury/ Illness/ Impairment	A070504	Premature Discharge of Battery	21/12/2023	21/12/2023
							F1903	Device Explantation	A070504	Premature Discharge of Battery	21/12/2023	21/12/2023
14658	2023	All other reportable incidents	47265	Dual-chamber implantable pacemaker, rate-responsive	CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A070101	Failure to Capture		
							F26	No Health Consequences or Impact	A070101	Failure to Capture		
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A070101	Failure to Capture		
							F26	No Health Consequences or Impact	A070101	Failure to Capture		
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A070101	Failure to Capture		
							F26	No Health Consequences or Impact	A070101	Failure to Capture		
14659	2023	All other reportable incidents	47267	Single-chamber implantable pacemaker, rate-responsive	CT141	Electrodes/Leads and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT145	Cardiovascular devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT233	Surgical	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT2406	Implantable	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT2407	Active implantable	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT2727	Active implant devices/ancillaries	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT2750	Implantable cardiac electrical stimulators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT2985	Active	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT3001	Implantable metal	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT3002	Implantable synthetic polymer	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT3006	Implantable inorganic material	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT314	Electrically powered	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT321	Long-term surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT449	Generators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		

						F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT981	Single use	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT983	Surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
14660	2023	All other reportable incidents	37265	Dual-chamber implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E1032	Vomiting	A070102	High Capture Threshold		
									A070909	Oversensing		
									A0712	Pacing Problem		
									A090801	Signal Artifact/Noise		
									A1301	Failure to Read Input Signal		
							F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A070102	High Capture Threshold		
									A070909	Oversensing		
									A0712	Pacing Problem		
									A090801	Signal Artifact/Noise		
									A1301	Failure to Read Input Signal		
							F2203	Imaging Required	A070102	High Capture Threshold		
									A070909	Oversensing		
									A0712	Pacing Problem		
									A090801	Signal Artifact/Noise		
									A1301	Failure to Read Input Signal		
14661	2023		35852	Single-chamber implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	F05	Delay to Treatment/ Therapy	A0203	Defective Device		
14662	2023	All other reportable incidents	47263	Cardiac resynchronization therapy implantable pacemaker	CT1685	Cardiac pacemakers	E2103	Device Overstimulation of Tissue	A0908	Incorrect, Inadequate or Imprecise Result or Readings	20/09/2023	20/09/2023
							F08	Hospitalization or Prolonged Hospitalization	A0908	Incorrect, Inadequate or Imprecise Result or Readings	20/09/2023	20/09/2023
							F15	Recognized Device or Procedural Complication	A0908	Incorrect, Inadequate or Imprecise Result or Readings	20/09/2023	20/09/2023
							F1905	Device Revision or Replacement	A0908	Incorrect, Inadequate or Imprecise Result or Readings	20/09/2023	20/09/2023
					CT257	Implantable cardiac pacemakers	E2103	Device Overstimulation of Tissue	A0908	Incorrect, Inadequate or Imprecise Result or Readings	20/09/2023	20/09/2023
							F08	Hospitalization or Prolonged Hospitalization	A0908	Incorrect, Inadequate or Imprecise Result or Readings	20/09/2023	20/09/2023
							F15	Recognized Device or Procedural Complication	A0908	Incorrect, Inadequate or Imprecise Result or Readings	20/09/2023	20/09/2023
							F1905	Device Revision or Replacement	A0908	Incorrect, Inadequate or Imprecise Result or Readings	20/09/2023	20/09/2023
					CT2866	Implantable cardiac electrical stimulators	E2103	Device Overstimulation of Tissue	A0908	Incorrect, Inadequate or Imprecise Result or Readings	20/09/2023	20/09/2023
							F08	Hospitalization or Prolonged Hospitalization	A0908	Incorrect, Inadequate or Imprecise Result or Readings	20/09/2023	20/09/2023
							F15	Recognized Device or Procedural Complication	A0908	Incorrect, Inadequate or Imprecise Result or Readings	20/09/2023	20/09/2023
							F1905	Device Revision or Replacement	A0908	Incorrect, Inadequate or Imprecise Result or Readings	20/09/2023	20/09/2023
14663	2023		47265	Dual-chamber implantable pacemaker, rate-responsive	CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	18/12/2023	18/12/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	18/12/2023	18/12/2023
							F1903	Device Explantation	A24	No Apparent Device Problem	18/12/2023	18/12/2023
							F1905	Device Revision or Replacement	A24	No Apparent Device Problem	18/12/2023	18/12/2023
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	18/12/2023	18/12/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	18/12/2023	18/12/2023
							F1903	Device Explantation	A24	No Apparent Device Problem	18/12/2023	18/12/2023
							F1905	Device Revision or Replacement	A24	No Apparent Device Problem	18/12/2023	18/12/2023
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	18/12/2023	18/12/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	18/12/2023	18/12/2023
							F1903	Device Explantation	A24	No Apparent Device Problem	18/12/2023	18/12/2023
							F1905	Device Revision or Replacement	A24	No Apparent Device Problem	18/12/2023	18/12/2023
14664	2023	All other reportable incidents	47265	Dual-chamber implantable pacemaker, rate-responsive	CT141	Electrodes/Leads and associated devices	E0112	Dizziness	A0712	Pacing Problem		
							E0601	Arrhythmia	A0712	Pacing Problem		
							F11	Non-Serious Injury/ Illness/ Impairment	A0712	Pacing Problem		
					CT145	Cardiovascular devices	E0112	Dizziness	A0712	Pacing Problem		
							E0601	Arrhythmia	A0712	Pacing Problem		
							F11	Non-Serious Injury/ Illness/ Impairment	A0712	Pacing Problem		
					CT1684	Cardiac pacemakers and associated devices	E0112	Dizziness	A0712	Pacing Problem		
							E0601	Arrhythmia	A0712	Pacing Problem		
							F11	Non-Serious Injury/ Illness/ Impairment	A0712	Pacing Problem		
					CT1685	Cardiac pacemakers	E0112	Dizziness	A0712	Pacing Problem		
							E0601	Arrhythmia	A0712	Pacing Problem		
							F11	Non-Serious Injury/ Illness/ Impairment	A0712	Pacing Problem		
					CT233	Surgical	E0112	Dizziness	A0712	Pacing Problem		
							E0601	Arrhythmia	A0712	Pacing Problem		
							F11	Non-Serious Injury/ Illness/ Impairment	A0712	Pacing Problem		
					CT2406	Implantable	E0112	Dizziness	A0712	Pacing Problem		
							E0601	Arrhythmia	A0712	Pacing Problem		
							F11	Non-Serious Injury/ Illness/ Impairment	A0712	Pacing Problem		
					CT2407	Active implantable	E0112	Dizziness	A0712	Pacing Problem		
							E0601	Arrhythmia	A0712	Pacing Problem		
							F11	Non-Serious Injury/ Illness/ Impairment	A0712	Pacing Problem		
					CT257	Implantable cardiac pacemakers	E0112	Dizziness	A0712	Pacing Problem		
							E0601	Arrhythmia	A0712	Pacing Problem		
							F11	Non-Serious Injury/ Illness/ Impairment	A0712	Pacing Problem		
					CT2727	Active implant devices/ancillaries	E0112	Dizziness	A0712	Pacing Problem		
							E0601	Arrhythmia	A0712	Pacing Problem		
							F11	Non-Serious Injury/ Illness/ Impairment	A0712	Pacing Problem		
					CT2750	Implantable cardiac electrical stimulators and associated devices	E0112	Dizziness	A0712	Pacing Problem		
							E0601	Arrhythmia	A0712	Pacing Problem		
							F11	Non-Serious Injury/ Illness/ Impairment	A0712	Pacing Problem		
					CT2866	Implantable cardiac electrical stimulators	E0112	Dizziness	A0712	Pacing Problem		
							E0601	Arrhythmia	A0712	Pacing Problem		
							F11	Non-Serious Injury/ Illness/ Impairment	A0712	Pacing Problem		
					CT2985	Active	E0112	Dizziness	A0712	Pacing Problem		

							E0601	Arrhythmia	A0712	Pacing Problem		
							F11	Non-Serious Injury/ Illness/ Impairment	A0712	Pacing Problem		
					CT3001	Implantable metal	E0112	Dizziness	A0712	Pacing Problem		
							E0601	Arrhythmia	A0712	Pacing Problem		
							F11	Non-Serious Injury/ Illness/ Impairment	A0712	Pacing Problem		
					CT3002	Implantable synthetic polymer	E0112	Dizziness	A0712	Pacing Problem		
							E0601	Arrhythmia	A0712	Pacing Problem		
							F11	Non-Serious Injury/ Illness/ Impairment	A0712	Pacing Problem		
					CT3006	Implantable inorganic material	E0112	Dizziness	A0712	Pacing Problem		
							E0601	Arrhythmia	A0712	Pacing Problem		
							F11	Non-Serious Injury/ Illness/ Impairment	A0712	Pacing Problem		
					CT314	Electrically powered	E0112	Dizziness	A0712	Pacing Problem		
							E0601	Arrhythmia	A0712	Pacing Problem		
							F11	Non-Serious Injury/ Illness/ Impairment	A0712	Pacing Problem		
					CT321	Long-term surgical invasive	E0112	Dizziness	A0712	Pacing Problem		
							E0601	Arrhythmia	A0712	Pacing Problem		
							F11	Non-Serious Injury/ Illness/ Impairment	A0712	Pacing Problem		
					CT449	Generators and associated devices	E0112	Dizziness	A0712	Pacing Problem		
							E0601	Arrhythmia	A0712	Pacing Problem		
							F11	Non-Serious Injury/ Illness/ Impairment	A0712	Pacing Problem		
					CT981	Single use	E0112	Dizziness	A0712	Pacing Problem		
							E0601	Arrhythmia	A0712	Pacing Problem		
							F11	Non-Serious Injury/ Illness/ Impairment	A0712	Pacing Problem		
					CT983	Surgical invasive	E0112	Dizziness	A0712	Pacing Problem		
							E0601	Arrhythmia	A0712	Pacing Problem		
							F11	Non-Serious Injury/ Illness/ Impairment	A0712	Pacing Problem		
14665	2023	All other reportable incidents	17882	Physiologic-monitoring defibrillation system	CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A0719	Unexpected Shutdown		
							F27	Problem Identified During Non-Clinical Procedure	A0719	Unexpected Shutdown		
14666	2023	All other reportable incidents	47265	Dual-chamber implantable pacemaker, rate-responsive	CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	23/08/2023	23/08/2023
									A160602	Premature End-of-Life Indicator	23/08/2023	23/08/2023
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	23/08/2023	23/08/2023
									A160602	Premature End-of-Life Indicator	23/08/2023	23/08/2023
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	23/08/2023	23/08/2023
									A160602	Premature End-of-Life Indicator	23/08/2023	23/08/2023
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	23/08/2023	23/08/2023
									A160602	Premature End-of-Life Indicator	23/08/2023	23/08/2023
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	23/08/2023	23/08/2023
									A160602	Premature End-of-Life Indicator	23/08/2023	23/08/2023
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	23/08/2023	23/08/2023
									A160602	Premature End-of-Life Indicator	23/08/2023	23/08/2023
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	23/08/2023	23/08/2023
									A160602	Premature End-of-Life Indicator	23/08/2023	23/08/2023
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	23/08/2023	23/08/2023
									A160602	Premature End-of-Life Indicator	23/08/2023	23/08/2023
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	23/08/2023	23/08/2023
									A160602	Premature End-of-Life Indicator	23/08/2023	23/08/2023
14667	2024	All other reportable incidents	35852	Single-chamber implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	12/12/2023	12/12/2023
									A0908	Incorrect, Inadequate or Imprecise Result or Readings	12/12/2023	12/12/2023
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	12/12/2023	12/12/2023
									A0908	Incorrect, Inadequate or Imprecise Result or Readings	12/12/2023	12/12/2023
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	12/12/2023	12/12/2023
									A0908	Incorrect, Inadequate or Imprecise Result or Readings	12/12/2023	12/12/2023
14668	2024	All other reportable incidents	47267	Single-chamber implantable pacemaker, rate-responsive	CT141	Electrodes/Leads and associated devices	E0717	Dyspnea	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
							E2312	Fatigue	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
							F24	Insufficient Health Impact Information	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
					CT145	Cardiovascular devices	E0717	Dyspnea	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
							E2312	Fatigue	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
							F24	Insufficient Health Impact Information	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
					CT1684	Cardiac pacemakers and associated devices	E0717	Dyspnea	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
							E2312	Fatigue	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
							F24	Insufficient Health Impact Information	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
					CT1685	Cardiac pacemakers	E0717	Dyspnea	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
							E2312	Fatigue	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
							F24	Insufficient Health Impact Information	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
					CT233	Surgical	E0717	Dyspnea	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
							E2312	Fatigue	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
							F24	Insufficient Health Impact Information	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
					CT2406	Implantable	E0717	Dyspnea	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
							E2312	Fatigue	A0705	Battery Problem		

									A070504	Premature Discharge of Battery		
							F24	Insufficient Health Impact Information	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
					CT2407	Active implantable	E0717	Dyspnea	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
							E2312	Fatigue	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
							F24	Insufficient Health Impact Information	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
					CT257	Implantable cardiac pacemakers	E0717	Dyspnea	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
							E2312	Fatigue	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
							F24	Insufficient Health Impact Information	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
					CT2727	Active implant devices/ancillaries	E0717	Dyspnea	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
							E2312	Fatigue	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
							F24	Insufficient Health Impact Information	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
					CT2750	Implantable cardiac electrical stimulators and associated devices	E0717	Dyspnea	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
							E2312	Fatigue	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
							F24	Insufficient Health Impact Information	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
					CT2866	Implantable cardiac electrical stimulators	E0717	Dyspnea	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
							E2312	Fatigue	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
							F24	Insufficient Health Impact Information	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
					CT2985	Active	E0717	Dyspnea	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
							E2312	Fatigue	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
							F24	Insufficient Health Impact Information	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
					CT3001	Implantable metal	E0717	Dyspnea	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
							E2312	Fatigue	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
							F24	Insufficient Health Impact Information	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
					CT3002	Implantable synthetic polymer	E0717	Dyspnea	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
							E2312	Fatigue	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
							F24	Insufficient Health Impact Information	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
					CT3006	Implantable inorganic material	E0717	Dyspnea	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
							E2312	Fatigue	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
							F24	Insufficient Health Impact Information	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
					CT314	Electrically powered	E0717	Dyspnea	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
							E2312	Fatigue	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
							F24	Insufficient Health Impact Information	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
					CT321	Long-term surgical invasive	E0717	Dyspnea	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
							E2312	Fatigue	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
							F24	Insufficient Health Impact Information	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
					CT449	Generators and associated devices	E0717	Dyspnea	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
							E2312	Fatigue	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
							F24	Insufficient Health Impact Information	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
					CT981	Single use	E0717	Dyspnea	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
							E2312	Fatigue	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
							F24	Insufficient Health Impact Information	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
					CT983	Surgical invasive	E0717	Dyspnea	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
							E2312	Fatigue	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
							F24	Insufficient Health Impact Information	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
14669	2024	All other reportable incidents	17882	Physiologic-monitoring defibrillation system	CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070903	Failure to Analyze Signal		

									A070908	Failure to Sense		
							F26	No Health Consequences or Impact	A070903	Failure to Analyze Signal		
									A070908	Failure to Sense		
14670	2024	All other reportable incidents	35852	Single-chamber implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E2104	Electric Shock	A070909	Oversensing		
									A071302	Inappropriate/Inadequate Shock/Stimulation		
									A1301	Failure to Read Input Signal		
							F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A070909	Oversensing		
									A071302	Inappropriate/Inadequate Shock/Stimulation		
									A1301	Failure to Read Input Signal		
14671	2024	All other reportable incidents	35852	Single-chamber implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A070909	Oversensing		
									A0712	Pacing Problem		
									A090801	Signal Artifact/Noise		
							F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A070909	Oversensing		
									A0712	Pacing Problem		
									A090801	Signal Artifact/Noise		
							F1001	Absence of Treatment	A070909	Oversensing		
									A0712	Pacing Problem		
									A090801	Signal Artifact/Noise		
14672	2024	All other reportable incidents	17882	Physiologic-monitoring defibrillation system	CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A071302	Inappropriate/Inadequate Shock/Stimulation		
							F27	Problem Identified During Non-Clinical Procedure	A071302	Inappropriate/Inadequate Shock/Stimulation		
14673	2024	All other reportable incidents	47265	Dual-chamber implantable pacemaker, rate-responsive	CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
							F24	Insufficient Health Impact Information	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
							F24	Insufficient Health Impact Information	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
							F24	Insufficient Health Impact Information	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
14674	2024		47265	Dual-chamber implantable pacemaker, rate-responsive	CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	05/12/2023	05/12/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	05/12/2023	05/12/2023
							F1903	Device Explantation	A24	No Apparent Device Problem	05/12/2023	05/12/2023
							F1905	Device Revision or Replacement	A24	No Apparent Device Problem	05/12/2023	05/12/2023
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	05/12/2023	05/12/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	05/12/2023	05/12/2023
							F1903	Device Explantation	A24	No Apparent Device Problem	05/12/2023	05/12/2023
							F1905	Device Revision or Replacement	A24	No Apparent Device Problem	05/12/2023	05/12/2023
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	05/12/2023	05/12/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	05/12/2023	05/12/2023
							F1903	Device Explantation	A24	No Apparent Device Problem	05/12/2023	05/12/2023
							F1905	Device Revision or Replacement	A24	No Apparent Device Problem	05/12/2023	05/12/2023
14675	2024	All other reportable incidents	60789	Intracardiac pacemaker	CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A051201	Device Dislodged or Dislocated	26/07/2022	26/07/2022
									A070101	Failure to Capture	26/07/2022	26/07/2022
									A070910	Undersensing	26/07/2022	26/07/2022
									A23	Use of Device Problem	26/07/2022	26/07/2022
							F08	Hospitalization or Prolonged Hospitalization	A051201	Device Dislodged or Dislocated	26/07/2022	26/07/2022
									A070101	Failure to Capture	26/07/2022	26/07/2022
									A070910	Undersensing	26/07/2022	26/07/2022
									A23	Use of Device Problem	26/07/2022	26/07/2022
							F1905	Device Revision or Replacement	A051201	Device Dislodged or Dislocated	26/07/2022	26/07/2022
									A070101	Failure to Capture	26/07/2022	26/07/2022
									A070910	Undersensing	26/07/2022	26/07/2022
									A23	Use of Device Problem	26/07/2022	26/07/2022
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A051201	Device Dislodged or Dislocated	26/07/2022	26/07/2022
									A070101	Failure to Capture	26/07/2022	26/07/2022
									A070910	Undersensing	26/07/2022	26/07/2022
									A23	Use of Device Problem	26/07/2022	26/07/2022
							F08	Hospitalization or Prolonged Hospitalization	A051201	Device Dislodged or Dislocated	26/07/2022	26/07/2022
									A070101	Failure to Capture	26/07/2022	26/07/2022
									A070910	Undersensing	26/07/2022	26/07/2022
									A23	Use of Device Problem	26/07/2022	26/07/2022
							F1905	Device Revision or Replacement	A051201	Device Dislodged or Dislocated	26/07/2022	26/07/2022
									A070101	Failure to Capture	26/07/2022	26/07/2022
									A070910	Undersensing	26/07/2022	26/07/2022
									A23	Use of Device Problem	26/07/2022	26/07/2022
					CT2818	Catheterization/Cannulation/Fluid path devices	E2403	No Clinical Signs, Symptoms or Conditions	A051201	Device Dislodged or Dislocated	26/07/2022	26/07/2022
									A070101	Failure to Capture	26/07/2022	26/07/2022
									A070910	Undersensing	26/07/2022	26/07/2022
									A23	Use of Device Problem	26/07/2022	26/07/2022
							F08	Hospitalization or Prolonged Hospitalization	A051201	Device Dislodged or Dislocated	26/07/2022	26/07/2022
									A070101	Failure to Capture	26/07/2022	26/07/2022
									A070910	Undersensing	26/07/2022	26/07/2022
									A23	Use of Device Problem	26/07/2022	26/07/2022
							F1905	Device Revision or Replacement	A051201	Device Dislodged or Dislocated	26/07/2022	26/07/2022
									A070101	Failure to Capture	26/07/2022	26/07/2022
									A070910	Undersensing	26/07/2022	26/07/2022
									A23	Use of Device Problem	26/07/2022	26/07/2022
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A051201	Device Dislodged or Dislocated	26/07/2022	26/07/2022
									A070101	Failure to Capture	26/07/2022	26/07/2022
									A070910	Undersensing	26/07/2022	26/07/2022
									A23	Use of Device Problem	26/07/2022	26/07/2022
							F08	Hospitalization or Prolonged Hospitalization	A051201	Device Dislodged or Dislocated	26/07/2022	26/07/2022
									A070101	Failure to Capture	26/07/2022	26/07/2022
									A070910	Undersensing	26/07/2022	26/07/2022
									A23	Use of Device Problem	26/07/2022	26/07/2022

							F1905	Device Revision or Replacement	A051201	Device Dislodged or Dislocated	26/07/2022	26/07/2022
									A070101	Failure to Capture	26/07/2022	26/07/2022
									A070910	Undersensing	26/07/2022	26/07/2022
									A23	Use of Device Problem	26/07/2022	26/07/2022
14676	2024	All other reportable incidents	47205	Cardiac pulse generator programmer	CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A110401	Operating System Becomes Nonfunctional		
							F26	No Health Consequences or Impact	A110401	Operating System Becomes Nonfunctional		
14677	2024	All other reportable incidents	37265	Dual-chamber implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E060109	Tachycardia	A0712	Pacing Problem		
							F26	No Health Consequences or Impact	A0712	Pacing Problem		
14678	2024	All other reportable incidents	35852	Single-chamber implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	01/01/2024	01/01/2024
									A0908	Incorrect, Inadequate or Imprecise Result or Readings	01/01/2024	01/01/2024
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	01/01/2024	01/01/2024
									A0908	Incorrect, Inadequate or Imprecise Result or Readings	01/01/2024	01/01/2024
							F1903	Device Explantation	A070504	Premature Discharge of Battery	01/01/2024	01/01/2024
									A0908	Incorrect, Inadequate or Imprecise Result or Readings	01/01/2024	01/01/2024
14679	2024	All other reportable incidents	47263	Cardiac resynchronization therapy implantable pacemaker	CT1685	Cardiac pacemakers	E011903	Syncope/Fainting	A0711	Interrogation Problem	28/04/2023	
									A0712	Pacing Problem	28/04/2023	
							F1001	Absence of Treatment	A0711	Interrogation Problem	28/04/2023	
									A0712	Pacing Problem	28/04/2023	
					CT257	Implantable cardiac pacemakers	E011903	Syncope/Fainting	A0711	Interrogation Problem	28/04/2023	
									A0712	Pacing Problem	28/04/2023	
							F1001	Absence of Treatment	A0711	Interrogation Problem	28/04/2023	
									A0712	Pacing Problem	28/04/2023	
					CT2866	Implantable cardiac electrical stimulators	E011903	Syncope/Fainting	A0711	Interrogation Problem	28/04/2023	
									A0712	Pacing Problem	28/04/2023	
							F1001	Absence of Treatment	A0711	Interrogation Problem	28/04/2023	
									A0712	Pacing Problem	28/04/2023	
14680	2024	All other reportable incidents	35852	Single-chamber implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A072201	High Impedance		
							F26	No Health Consequences or Impact	A072201	High Impedance		
14681	2024	All other reportable incidents	47265	Dual-chamber implantable pacemaker, rate-responsive	CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
14682	2024	All other reportable incidents	35852	Single-chamber implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E2104	Electric Shock	A070909	Oversensing		
									A071302	Inappropriate/Inadequate Shock/Stimulation		
							F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A070909	Oversensing		
									A071302	Inappropriate/Inadequate Shock/Stimulation		
							F15	Recognized Device or Procedural Complication	A070909	Oversensing		
									A071302	Inappropriate/Inadequate Shock/Stimulation		
14683	2024	All other reportable incidents	35852	Single-chamber implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E2104	Electric Shock	A070909	Oversensing		
									A070910	Undersensing		
									A071302	Inappropriate/Inadequate Shock/Stimulation		
									A090801	Signal Artifact/Noise		
							F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A070909	Oversensing		
									A070910	Undersensing		
									A071302	Inappropriate/Inadequate Shock/Stimulation		
									A090801	Signal Artifact/Noise		
							F15	Recognized Device or Procedural Complication	A070909	Oversensing		
									A070910	Undersensing		
									A071302	Inappropriate/Inadequate Shock/Stimulation		
									A090801	Signal Artifact/Noise		
14684	2024		47265	Dual-chamber implantable pacemaker, rate-responsive	CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	21/11/2023	21/11/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	21/11/2023	21/11/2023
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	21/11/2023	21/11/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	21/11/2023	21/11/2023
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	21/11/2023	21/11/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	21/11/2023	21/11/2023
14685	2024		47265	Dual-chamber implantable pacemaker, rate-responsive	CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	09/11/2023	09/11/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	09/11/2023	09/11/2023
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	09/11/2023	09/11/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	09/11/2023	09/11/2023
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	09/11/2023	09/11/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	09/11/2023	09/11/2023
14686	2024	All other reportable incidents	17882	Physiologic-monitoring defibrillation system	CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F27	Problem Identified During Non-Clinical Procedure	A070504	Premature Discharge of Battery		
14687	2024	All other reportable incidents	47205	Cardiac pulse generator programmer	CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A09	Output Problem		
									A13	Communication or Transmission Problem		
							F23	Unexpected Medical Intervention	A09	Output Problem		
									A13	Communication or Transmission Problem		
14688	2024		35852	Single-chamber implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem		
							F26	No Health Consequences or Impact	A0705	Battery Problem		
14689	2024	All other reportable incidents	17882	Physiologic-monitoring defibrillation system	CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F27	Problem Identified During Non-Clinical Procedure	A070504	Premature Discharge of Battery		
14690	2024	All other reportable incidents	47263	Cardiac resynchronization therapy implantable pacemaker	CT1685	Cardiac pacemakers	E060104	Bradycardia	A070909	Oversensing		
									A0712	Pacing Problem		
									A072201	High Impedance		
							F1001	Absence of Treatment	A070909	Oversensing		
									A0712	Pacing Problem		
									A072201	High Impedance		
							F2203	Imaging Required	A070909	Oversensing		
									A0712	Pacing Problem		
									A072201	High Impedance		
							F2301	Additional Device Required	A070909	Oversensing		
									A0712	Pacing Problem		

									A072201	High Impedance		
					CT257	Implantable cardiac pacemakers	E060104	Bradycardia	A070909	Oversensing		
									A0712	Pacing Problem		
									A072201	High Impedance		
							F1001	Absence of Treatment	A070909	Oversensing		
									A0712	Pacing Problem		
									A072201	High Impedance		
							F2203	Imaging Required	A070909	Oversensing		
									A0712	Pacing Problem		
									A072201	High Impedance		
							F2301	Additional Device Required	A070909	Oversensing		
									A0712	Pacing Problem		
									A072201	High Impedance		
					CT2866	Implantable cardiac electrical stimulators	E060104	Bradycardia	A070909	Oversensing		
									A0712	Pacing Problem		
									A072201	High Impedance		
							F1001	Absence of Treatment	A070909	Oversensing		
									A0712	Pacing Problem		
									A072201	High Impedance		
							F2203	Imaging Required	A070909	Oversensing		
									A0712	Pacing Problem		
									A072201	High Impedance		
							F2301	Additional Device Required	A070909	Oversensing		
									A0712	Pacing Problem		
									A072201	High Impedance		
14691	2024	All other reportable incidents			CT257	Implantable cardiac pacemakers	E2104	Electric Shock	A040101	Fracture	04/01/2024	04/01/2024
									A070909	Oversensing	04/01/2024	04/01/2024
									A071302	Inappropriate/Inadequate Shock/Stimulation	04/01/2024	04/01/2024
							F08	Hospitalization or Prolonged Hospitalization	A040101	Fracture	04/01/2024	04/01/2024
									A070909	Oversensing	04/01/2024	04/01/2024
									A071302	Inappropriate/Inadequate Shock/Stimulation	04/01/2024	04/01/2024
							F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A040101	Fracture	04/01/2024	04/01/2024
									A070909	Oversensing	04/01/2024	04/01/2024
									A071302	Inappropriate/Inadequate Shock/Stimulation	04/01/2024	04/01/2024
							F15	Recognized Device or Procedural Complication	A040101	Fracture	04/01/2024	04/01/2024
									A070909	Oversensing	04/01/2024	04/01/2024
									A071302	Inappropriate/Inadequate Shock/Stimulation	04/01/2024	04/01/2024
							F1905	Device Revision or Replacement	A040101	Fracture	04/01/2024	04/01/2024
									A070909	Oversensing	04/01/2024	04/01/2024
									A071302	Inappropriate/Inadequate Shock/Stimulation	04/01/2024	04/01/2024
14692	2024	All other reportable incidents	37265	Dual-chamber implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A070101	Failure to Capture		
									A072201	High Impedance		
									A090801	Signal Artifact/Noise		
							F2203	Imaging Required	A070101	Failure to Capture		
									A072201	High Impedance		
									A090801	Signal Artifact/Noise		
14693	2024	All other reportable incidents	35852	Single-chamber implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E0402	Hypersensitivity/Allergic Reaction	A24	No Apparent Device Problem		
							E1707	Impaired Healing	A24	No Apparent Device Problem		
							F15	Recognized Device or Procedural Complication	A24	No Apparent Device Problem		
14694	2024	All other reportable incidents	47270	Cardiac resynchronization therapy implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A070909	Oversensing		
									A0712	Pacing Problem		
									A090801	Signal Artifact/Noise		
							F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A070909	Oversensing		
									A0712	Pacing Problem		
									A090801	Signal Artifact/Noise		
14695	2024	All other reportable incidents	47263	Cardiac resynchronization therapy implantable pacemaker	CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
14696	2024		47265	Dual-chamber implantable pacemaker, rate-responsive	CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	15/12/2023	15/12/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	15/12/2023	15/12/2023
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	15/12/2023	15/12/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	15/12/2023	15/12/2023
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	15/12/2023	15/12/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	15/12/2023	15/12/2023
14697	2024	All other reportable incidents	17882	Physiologic-monitoring defibrillation system	CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A071301	Failure to Deliver Shock/Stimulation		
							F26	No Health Consequences or Impact	A071301	Failure to Deliver Shock/Stimulation		
14698	2024	All other reportable incidents	47265	Dual-chamber implantable pacemaker, rate-responsive	CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem	18/01/2024	18/01/2024
									A070504	Premature Discharge of Battery	18/01/2024	18/01/2024
							F08	Hospitalization or Prolonged Hospitalization	A0705	Battery Problem	18/01/2024	18/01/2024
									A070504	Premature Discharge of Battery	18/01/2024	18/01/2024
							F1905	Device Revision or Replacement	A0705	Battery Problem	18/01/2024	18/01/2024
									A070504	Premature Discharge of Battery	18/01/2024	18/01/2024
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem	18/01/2024	18/01/2024
									A070504	Premature Discharge of Battery	18/01/2024	18/01/2024
							F08	Hospitalization or Prolonged Hospitalization	A0705	Battery Problem	18/01/2024	18/01/2024
									A070504	Premature Discharge of Battery	18/01/2024	18/01/2024
							F1905	Device Revision or Replacement	A0705	Battery Problem	18/01/2024	18/01/2024
									A070504	Premature Discharge of Battery	18/01/2024	18/01/2024
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem	18/01/2024	18/01/2024

									A070504	Premature Discharge of Battery	18/01/2024	18/01/2024
							F08	Hospitalization or Prolonged Hospitalization	A0705	Battery Problem	18/01/2024	18/01/2024
									A070504	Premature Discharge of Battery	18/01/2024	18/01/2024
							F1905	Device Revision or Replacement	A0705	Battery Problem	18/01/2024	18/01/2024
									A070504	Premature Discharge of Battery	18/01/2024	18/01/2024
14699	2024	All other reportable incidents	35852	Single-chamber implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	18/01/2024	18/01/2024
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	18/01/2024	18/01/2024
							F12	Serious Injury/ Illness/ Impairment	A070504	Premature Discharge of Battery	18/01/2024	18/01/2024
							F1903	Device Explantation	A070504	Premature Discharge of Battery	18/01/2024	18/01/2024
14700	2024	All other reportable incidents	35852	Single-chamber implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E060110	Ventricular Fibrillation	A070910	Undersensing	01/01/2024	01/01/2024
									A071202	Inaccurate Synchronization	01/01/2024	01/01/2024
							F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A070910	Undersensing	01/01/2024	01/01/2024
									A071202	Inaccurate Synchronization	01/01/2024	01/01/2024
							F1203	Life Threatening Illness or Injury	A070910	Undersensing	01/01/2024	01/01/2024
									A071202	Inaccurate Synchronization	01/01/2024	01/01/2024
							F15	Recognized Device or Procedural Complication	A070910	Undersensing	01/01/2024	01/01/2024
									A071202	Inaccurate Synchronization	01/01/2024	01/01/2024
14701	2024	All other reportable incidents	47263	Cardiac resynchronization therapy implantable pacemaker	CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A0908	Incorrect, Inadequate or Imprecise Result or Readings	18/01/2024	18/01/2024
							F08	Hospitalization or Prolonged Hospitalization	A0908	Incorrect, Inadequate or Imprecise Result or Readings	18/01/2024	18/01/2024
							F1905	Device Revision or Replacement	A0908	Incorrect, Inadequate or Imprecise Result or Readings	18/01/2024	18/01/2024
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A0908	Incorrect, Inadequate or Imprecise Result or Readings	18/01/2024	18/01/2024
							F08	Hospitalization or Prolonged Hospitalization	A0908	Incorrect, Inadequate or Imprecise Result or Readings	18/01/2024	18/01/2024
							F1905	Device Revision or Replacement	A0908	Incorrect, Inadequate or Imprecise Result or Readings	18/01/2024	18/01/2024
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A0908	Incorrect, Inadequate or Imprecise Result or Readings	18/01/2024	18/01/2024
							F08	Hospitalization or Prolonged Hospitalization	A0908	Incorrect, Inadequate or Imprecise Result or Readings	18/01/2024	18/01/2024
							F1905	Device Revision or Replacement	A0908	Incorrect, Inadequate or Imprecise Result or Readings	18/01/2024	18/01/2024
14702	2024	All other reportable incidents	37265	Dual-chamber implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A070909	Oversensing		
									A071204	No Pacing		
							F22	Unexpected Diagnostic Intervention	A070909	Oversensing		
									A071204	No Pacing		
14703	2024		47265	Dual-chamber implantable pacemaker, rate-responsive	CT141	Electrodes/Leads and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	14/12/2023	14/12/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	14/12/2023	14/12/2023
					CT145	Cardiovascular devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	14/12/2023	14/12/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	14/12/2023	14/12/2023
					CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	14/12/2023	14/12/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	14/12/2023	14/12/2023
					CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	14/12/2023	14/12/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	14/12/2023	14/12/2023
					CT233	Surgical	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	14/12/2023	14/12/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	14/12/2023	14/12/2023
					CT2406	Implantable	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	14/12/2023	14/12/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	14/12/2023	14/12/2023
					CT2407	Active implantable	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	14/12/2023	14/12/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	14/12/2023	14/12/2023
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	14/12/2023	14/12/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	14/12/2023	14/12/2023
					CT2727	Active implant devices/ancillaries	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	14/12/2023	14/12/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	14/12/2023	14/12/2023
					CT2750	Implantable cardiac electrical stimulators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	14/12/2023	14/12/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	14/12/2023	14/12/2023
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	14/12/2023	14/12/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	14/12/2023	14/12/2023
					CT2985	Active	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	14/12/2023	14/12/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	14/12/2023	14/12/2023
					CT3001	Implantable metal	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	14/12/2023	14/12/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	14/12/2023	14/12/2023
					CT3002	Implantable synthetic polymer	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	14/12/2023	14/12/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	14/12/2023	14/12/2023
					CT3006	Implantable inorganic material	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	14/12/2023	14/12/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	14/12/2023	14/12/2023
					CT314	Electrically powered	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	14/12/2023	14/12/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	14/12/2023	14/12/2023
					CT321	Long-term surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	14/12/2023	14/12/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	14/12/2023	14/12/2023
					CT449	Generators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	14/12/2023	14/12/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	14/12/2023	14/12/2023
					CT981	Single use	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	14/12/2023	14/12/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	14/12/2023	14/12/2023
					CT983	Surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	14/12/2023	14/12/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	14/12/2023	14/12/2023
14704	2024	All other reportable incidents	47265	Dual-chamber implantable pacemaker, rate-responsive	CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A071102	Failure to Interrogate	09/01/2024	09/01/2024
							F1901	Additional Surgery	A071102	Failure to Interrogate	09/01/2024	09/01/2024
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A071102	Failure to Interrogate	09/01/2024	09/01/2024
							F1901	Additional Surgery	A071102	Failure to Interrogate	09/01/2024	09/01/2024
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A071102	Failure to Interrogate	09/01/2024	09/01/2024
							F1901	Additional Surgery	A071102	Failure to Interrogate	09/01/2024	09/01/2024
14705	2024		47265	Dual-chamber implantable pacemaker, rate-responsive	CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	17/04/2023	17/04/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	17/04/2023	17/04/2023
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	17/04/2023	17/04/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	17/04/2023	17/04/2023
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	17/04/2023	17/04/2023

							F1901	Additional Surgery	A24	No Apparent Device Problem	17/04/2023	17/04/2023
14706	2024	All other reportable incidents	17882	Physiologic-monitoring defibrillation system	CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A0712	Pacing Problem		
							F27	Problem Identified During Non-Clinical Procedure	A0712	Pacing Problem		
14707	2024	All other reportable incidents	17882	Physiologic-monitoring defibrillation system	CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem		
							F27	Problem Identified During Non-Clinical Procedure	A0705	Battery Problem		
14708	2024		47265	Dual-chamber implantable pacemaker, rate-responsive	CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	28/11/2023	28/11/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	28/11/2023	28/11/2023
							F1903	Device Explantation	A24	No Apparent Device Problem	28/11/2023	28/11/2023
							F1905	Device Revision or Replacement	A24	No Apparent Device Problem	28/11/2023	28/11/2023
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	28/11/2023	28/11/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	28/11/2023	28/11/2023
							F1903	Device Explantation	A24	No Apparent Device Problem	28/11/2023	28/11/2023
							F1905	Device Revision or Replacement	A24	No Apparent Device Problem	28/11/2023	28/11/2023
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	28/11/2023	28/11/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	28/11/2023	28/11/2023
							F1903	Device Explantation	A24	No Apparent Device Problem	28/11/2023	28/11/2023
							F1905	Device Revision or Replacement	A24	No Apparent Device Problem	28/11/2023	28/11/2023
14709	2024		47265	Dual-chamber implantable pacemaker, rate-responsive	CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	15/11/2023	15/11/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	15/11/2023	15/11/2023
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	15/11/2023	15/11/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	15/11/2023	15/11/2023
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	15/11/2023	15/11/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	15/11/2023	15/11/2023
14710	2024		47265	Dual-chamber implantable pacemaker, rate-responsive	CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	08/11/2023	08/11/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	08/11/2023	08/11/2023
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	08/11/2023	08/11/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	08/11/2023	08/11/2023
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	08/11/2023	08/11/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	08/11/2023	08/11/2023
14711	2024	All other reportable incidents	35852	Single-chamber implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A072201	High Impedance		
							F26	No Health Consequences or Impact	A072201	High Impedance		
14712	2024	All other reportable incidents	36241	Pacemaker/defibrillator lead	CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A072201	High Impedance		
							F26	No Health Consequences or Impact	A072201	High Impedance		
14713	2024	All other reportable incidents	35852	Single-chamber implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	16/03/2023	16/03/2023
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	16/03/2023	16/03/2023
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	16/03/2023	16/03/2023
14714	2024	Death	47265	Dual-chamber implantable pacemaker, rate-responsive	CT1685	Cardiac pacemakers	E011903	Syncope/Fainting	A071102	Failure to Interrogate	12/01/2024	12/01/2024
									A071204	No Pacing	12/01/2024	12/01/2024
							E0601	Arrhythmia	A071102	Failure to Interrogate	12/01/2024	12/01/2024
									A071204	No Pacing	12/01/2024	12/01/2024
							E0602	Cardiac Arrest	A071102	Failure to Interrogate	12/01/2024	12/01/2024
									A071204	No Pacing	12/01/2024	12/01/2024
							E0742	Respiratory Failure	A071102	Failure to Interrogate	12/01/2024	12/01/2024
									A071204	No Pacing	12/01/2024	12/01/2024
							E2402	Appropriate Clinical Signs, Symptoms and Conditions Term/Code not Available	A071102	Failure to Interrogate	12/01/2024	12/01/2024
									A071204	No Pacing	12/01/2024	12/01/2024
							F02	Death	A071102	Failure to Interrogate	12/01/2024	12/01/2024
									A071204	No Pacing	12/01/2024	12/01/2024
					CT257	Implantable cardiac pacemakers	E011903	Syncope/Fainting	A071102	Failure to Interrogate	12/01/2024	12/01/2024
									A071204	No Pacing	12/01/2024	12/01/2024
							E0601	Arrhythmia	A071102	Failure to Interrogate	12/01/2024	12/01/2024
									A071204	No Pacing	12/01/2024	12/01/2024
							E0602	Cardiac Arrest	A071102	Failure to Interrogate	12/01/2024	12/01/2024
									A071204	No Pacing	12/01/2024	12/01/2024
							E0742	Respiratory Failure	A071102	Failure to Interrogate	12/01/2024	12/01/2024
									A071204	No Pacing	12/01/2024	12/01/2024
							E2402	Appropriate Clinical Signs, Symptoms and Conditions Term/Code not Available	A071102	Failure to Interrogate	12/01/2024	12/01/2024
									A071204	No Pacing	12/01/2024	12/01/2024
							F02	Death	A071102	Failure to Interrogate	12/01/2024	12/01/2024
									A071204	No Pacing	12/01/2024	12/01/2024
					CT2866	Implantable cardiac electrical stimulators	E011903	Syncope/Fainting	A071102	Failure to Interrogate	12/01/2024	12/01/2024
									A071204	No Pacing	12/01/2024	12/01/2024
							E0601	Arrhythmia	A071102	Failure to Interrogate	12/01/2024	12/01/2024
									A071204	No Pacing	12/01/2024	12/01/2024
							E0602	Cardiac Arrest	A071102	Failure to Interrogate	12/01/2024	12/01/2024
									A071204	No Pacing	12/01/2024	12/01/2024
							E0742	Respiratory Failure	A071102	Failure to Interrogate	12/01/2024	12/01/2024
									A071204	No Pacing	12/01/2024	12/01/2024
							E2402	Appropriate Clinical Signs, Symptoms and Conditions Term/Code not Available	A071102	Failure to Interrogate	12/01/2024	12/01/2024
									A071204	No Pacing	12/01/2024	12/01/2024
							F02	Death	A071102	Failure to Interrogate	12/01/2024	12/01/2024
									A071204	No Pacing	12/01/2024	12/01/2024
14715	2024	All other reportable incidents	47263	Cardiac resynchronization therapy implantable pacemaker	CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	10/01/2024	10/01/2024
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	10/01/2024	10/01/2024
							F12	Serious Injury/ Illness/ Impairment	A070504	Premature Discharge of Battery	10/01/2024	10/01/2024
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	10/01/2024	10/01/2024
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	10/01/2024	10/01/2024
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	10/01/2024	10/01/2024
							F12	Serious Injury/ Illness/ Impairment	A070504	Premature Discharge of Battery	10/01/2024	10/01/2024
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	10/01/2024	10/01/2024
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	10/01/2024	10/01/2024
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	10/01/2024	10/01/2024
							F12	Serious Injury/ Illness/ Impairment	A070504	Premature Discharge of Battery	10/01/2024	10/01/2024
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	10/01/2024	10/01/2024
14716	2024		47265	Dual-chamber implantable pacemaker, rate-responsive	CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A090801	Signal Artifact/Noise	08/01/2024	08/01/2024
									A1205	Loose or Intermittent Connection	08/01/2024	08/01/2024
							F1901	Additional Surgery	A090801	Signal Artifact/Noise	08/01/2024	08/01/2024
									A1205	Loose or Intermittent Connection	08/01/2024	08/01/2024

					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A090801	Signal Artifact/Noise	08/01/2024	08/01/2024
									A1205	Loose or Intermittent Connection	08/01/2024	08/01/2024
							F1901	Additional Surgery	A090801	Signal Artifact/Noise	08/01/2024	08/01/2024
									A1205	Loose or Intermittent Connection	08/01/2024	08/01/2024
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A090801	Signal Artifact/Noise	08/01/2024	08/01/2024
									A1205	Loose or Intermittent Connection	08/01/2024	08/01/2024
							F1901	Additional Surgery	A090801	Signal Artifact/Noise	08/01/2024	08/01/2024
									A1205	Loose or Intermittent Connection	08/01/2024	08/01/2024
14717	2024	All other reportable incidents	17882	Physiologic-monitoring defibrillation system	CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A0721	Circuit Failure		
							F27	Problem Identified During Non-Clinical Procedure	A0721	Circuit Failure		
14718	2024		47265	Dual-chamber implantable pacemaker, rate-responsive	CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	08/11/2023	08/11/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	08/11/2023	08/11/2023
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	08/11/2023	08/11/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	08/11/2023	08/11/2023
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	08/11/2023	08/11/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	08/11/2023	08/11/2023
14719	2024	All other reportable incidents	47265	Dual-chamber implantable pacemaker, rate-responsive	CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	17/11/2023	17/11/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	17/11/2023	17/11/2023
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	17/11/2023	17/11/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	17/11/2023	17/11/2023
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	17/11/2023	17/11/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	17/11/2023	17/11/2023
14720	2024	All other reportable incidents	47265	Dual-chamber implantable pacemaker, rate-responsive	CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
14721	2024	All other reportable incidents	47267	Single-chamber implantable pacemaker, rate-responsive	CT1685	Cardiac pacemakers	E0112	Dizziness	A070101	Failure to Capture	23/11/2023	23/11/2023
									A070102	High Capture Threshold	23/11/2023	23/11/2023
									A070909	Oversensing	23/11/2023	23/11/2023
									A0712	Pacing Problem	23/11/2023	23/11/2023
							E0119	Loss of Consciousness	A070101	Failure to Capture	23/11/2023	23/11/2023
									A070102	High Capture Threshold	23/11/2023	23/11/2023
									A070909	Oversensing	23/11/2023	23/11/2023
									A0712	Pacing Problem	23/11/2023	23/11/2023
							E011903	Syncope/Fainting	A070101	Failure to Capture	23/11/2023	23/11/2023
									A070102	High Capture Threshold	23/11/2023	23/11/2023
									A070909	Oversensing	23/11/2023	23/11/2023
									A0712	Pacing Problem	23/11/2023	23/11/2023
							E2007	Fall	A070101	Failure to Capture	23/11/2023	23/11/2023
									A070102	High Capture Threshold	23/11/2023	23/11/2023
									A070909	Oversensing	23/11/2023	23/11/2023
									A0712	Pacing Problem	23/11/2023	23/11/2023
							F08	Hospitalization or Prolonged Hospitalization	A070101	Failure to Capture	23/11/2023	23/11/2023
									A070102	High Capture Threshold	23/11/2023	23/11/2023
									A070909	Oversensing	23/11/2023	23/11/2023
									A0712	Pacing Problem	23/11/2023	23/11/2023
							F1001	Absence of Treatment	A070101	Failure to Capture	23/11/2023	23/11/2023
									A070102	High Capture Threshold	23/11/2023	23/11/2023
									A070909	Oversensing	23/11/2023	23/11/2023
									A0712	Pacing Problem	23/11/2023	23/11/2023
							F1905	Device Revision or Replacement	A070101	Failure to Capture	23/11/2023	23/11/2023
									A070102	High Capture Threshold	23/11/2023	23/11/2023
									A070909	Oversensing	23/11/2023	23/11/2023
									A0712	Pacing Problem	23/11/2023	23/11/2023
							F2203	Imaging Required	A070101	Failure to Capture	23/11/2023	23/11/2023
									A070102	High Capture Threshold	23/11/2023	23/11/2023
									A070909	Oversensing	23/11/2023	23/11/2023
									A0712	Pacing Problem	23/11/2023	23/11/2023
					CT257	Implantable cardiac pacemakers	E0112	Dizziness	A070101	Failure to Capture	23/11/2023	23/11/2023
									A070102	High Capture Threshold	23/11/2023	23/11/2023
									A070909	Oversensing	23/11/2023	23/11/2023
									A0712	Pacing Problem	23/11/2023	23/11/2023
							E0119	Loss of Consciousness	A070101	Failure to Capture	23/11/2023	23/11/2023
									A070102	High Capture Threshold	23/11/2023	23/11/2023
									A070909	Oversensing	23/11/2023	23/11/2023
									A0712	Pacing Problem	23/11/2023	23/11/2023
							E011903	Syncope/Fainting	A070101	Failure to Capture	23/11/2023	23/11/2023
									A070102	High Capture Threshold	23/11/2023	23/11/2023
									A070909	Oversensing	23/11/2023	23/11/2023
									A0712	Pacing Problem	23/11/2023	23/11/2023
							E2007	Fall	A070101	Failure to Capture	23/11/2023	23/11/2023
									A070102	High Capture Threshold	23/11/2023	23/11/2023
									A070909	Oversensing	23/11/2023	23/11/2023
									A0712	Pacing Problem	23/11/2023	23/11/2023
							F08	Hospitalization or Prolonged Hospitalization	A070101	Failure to Capture	23/11/2023	23/11/2023
									A070102	High Capture Threshold	23/11/2023	23/11/2023
									A070909	Oversensing	23/11/2023	23/11/2023
									A0712	Pacing Problem	23/11/2023	23/11/2023
							F1001	Absence of Treatment	A070101	Failure to Capture	23/11/2023	23/11/2023
									A070102	High Capture Threshold	23/11/2023	23/11/2023
									A070909	Oversensing	23/11/2023	23/11/2023
									A0712	Pacing Problem	23/11/2023	23/11/2023
							F1905	Device Revision or Replacement	A070101	Failure to Capture	23/11/2023	23/11/2023
									A070102	High Capture Threshold	23/11/2023	23/11/2023
									A070909	Oversensing	23/11/2023	23/11/2023
									A0712	Pacing Problem	23/11/2023	23/11/2023

							F2203	Imaging Required	A070101	Failure to Capture	23/11/2023	23/11/2023
									A070102	High Capture Threshold	23/11/2023	23/11/2023
									A070909	Oversensing	23/11/2023	23/11/2023
									A0712	Pacing Problem	23/11/2023	23/11/2023
					CT2866	Implantable cardiac electrical stimulators	E0112	Dizziness	A070101	Failure to Capture	23/11/2023	23/11/2023
									A070102	High Capture Threshold	23/11/2023	23/11/2023
									A070909	Oversensing	23/11/2023	23/11/2023
									A0712	Pacing Problem	23/11/2023	23/11/2023
							E0119	Loss of Consciousness	A070101	Failure to Capture	23/11/2023	23/11/2023
									A070102	High Capture Threshold	23/11/2023	23/11/2023
									A070909	Oversensing	23/11/2023	23/11/2023
									A0712	Pacing Problem	23/11/2023	23/11/2023
							E011903	Syncope/Fainting	A070101	Failure to Capture	23/11/2023	23/11/2023
									A070102	High Capture Threshold	23/11/2023	23/11/2023
									A070909	Oversensing	23/11/2023	23/11/2023
									A0712	Pacing Problem	23/11/2023	23/11/2023
							E2007	Fall	A070101	Failure to Capture	23/11/2023	23/11/2023
									A070102	High Capture Threshold	23/11/2023	23/11/2023
									A070909	Oversensing	23/11/2023	23/11/2023
									A0712	Pacing Problem	23/11/2023	23/11/2023
							F08	Hospitalization or Prolonged Hospitalization	A070101	Failure to Capture	23/11/2023	23/11/2023
									A070102	High Capture Threshold	23/11/2023	23/11/2023
									A070909	Oversensing	23/11/2023	23/11/2023
									A0712	Pacing Problem	23/11/2023	23/11/2023
							F1001	Absence of Treatment	A070101	Failure to Capture	23/11/2023	23/11/2023
									A070102	High Capture Threshold	23/11/2023	23/11/2023
									A070909	Oversensing	23/11/2023	23/11/2023
									A0712	Pacing Problem	23/11/2023	23/11/2023
							F1905	Device Revision or Replacement	A070101	Failure to Capture	23/11/2023	23/11/2023
									A070102	High Capture Threshold	23/11/2023	23/11/2023
									A070909	Oversensing	23/11/2023	23/11/2023
									A0712	Pacing Problem	23/11/2023	23/11/2023
							F2203	Imaging Required	A070101	Failure to Capture	23/11/2023	23/11/2023
									A070102	High Capture Threshold	23/11/2023	23/11/2023
									A070909	Oversensing	23/11/2023	23/11/2023
									A0712	Pacing Problem	23/11/2023	23/11/2023
14722	2024	All other reportable incidents	35852	Single-chamber implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	06/03/2024	06/03/2024
									A0908	Incorrect, Inadequate or Imprecise Result or Readings	06/03/2024	06/03/2024
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	06/03/2024	06/03/2024
									A0908	Incorrect, Inadequate or Imprecise Result or Readings	06/03/2024	06/03/2024
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	06/03/2024	06/03/2024
									A0908	Incorrect, Inadequate or Imprecise Result or Readings	06/03/2024	06/03/2024
14723	2024		35852	Single-chamber implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A070909	Oversensing		
									A071302	Inappropriate/Inadequate Shock/Stimulation		
							F12	Serious Injury/ Illness/ Impairment	A070909	Oversensing		
									A071302	Inappropriate/Inadequate Shock/Stimulation		
14724	2024	All other reportable incidents	47263	Cardiac resynchronization therapy implantable pacemaker	CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A0908	Incorrect, Inadequate or Imprecise Result or Readings	09/01/2024	09/01/2024
							F08	Hospitalization or Prolonged Hospitalization	A0908	Incorrect, Inadequate or Imprecise Result or Readings	09/01/2024	09/01/2024
							F1905	Device Revision or Replacement	A0908	Incorrect, Inadequate or Imprecise Result or Readings	09/01/2024	09/01/2024
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A0908	Incorrect, Inadequate or Imprecise Result or Readings	09/01/2024	09/01/2024
							F08	Hospitalization or Prolonged Hospitalization	A0908	Incorrect, Inadequate or Imprecise Result or Readings	09/01/2024	09/01/2024
							F1905	Device Revision or Replacement	A0908	Incorrect, Inadequate or Imprecise Result or Readings	09/01/2024	09/01/2024
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A0908	Incorrect, Inadequate or Imprecise Result or Readings	09/01/2024	09/01/2024
							F08	Hospitalization or Prolonged Hospitalization	A0908	Incorrect, Inadequate or Imprecise Result or Readings	09/01/2024	09/01/2024
							F1905	Device Revision or Replacement	A0908	Incorrect, Inadequate or Imprecise Result or Readings	09/01/2024	09/01/2024
14725	2024	All other reportable incidents	37265	Dual-chamber implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A070910	Undersensing		
									A0712	Pacing Problem		
							F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A070910	Undersensing		
									A0712	Pacing Problem		
14726	2024		47265	Dual-chamber implantable pacemaker, rate-responsive	CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A0203	Defective Device		
									A07	Electrical /Electronic Property Problem		
							F1903	Device Explantation	A0203	Defective Device		
									A07	Electrical /Electronic Property Problem		
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A0203	Defective Device		
									A07	Electrical /Electronic Property Problem		
							F1903	Device Explantation	A0203	Defective Device		
									A07	Electrical /Electronic Property Problem		
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A0203	Defective Device		
									A07	Electrical /Electronic Property Problem		
							F1903	Device Explantation	A0203	Defective Device		
									A07	Electrical /Electronic Property Problem		
14727	2024	Unanticipated serious deterioration in state of health	47267	Single-chamber implantable pacemaker, rate-responsive	CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A070101	Failure to Capture	08/01/2024	
									A070909	Oversensing	08/01/2024	
									A090801	Signal Artifact/Noise	08/01/2024	
							F1903	Device Explantation	A070101	Failure to Capture	08/01/2024	
									A070909	Oversensing	08/01/2024	
									A090801	Signal Artifact/Noise	08/01/2024	
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A070101	Failure to Capture	08/01/2024	
									A070909	Oversensing	08/01/2024	

									A090801	Signal Artifact/Noise	08/01/2024	
							F1903	Device Explantation	A070101	Failure to Capture	08/01/2024	
									A070909	Oversensing	08/01/2024	
									A090801	Signal Artifact/Noise	08/01/2024	
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A070101	Failure to Capture	08/01/2024	
									A070909	Oversensing	08/01/2024	
									A090801	Signal Artifact/Noise	08/01/2024	
							F1903	Device Explantation	A070101	Failure to Capture	08/01/2024	
									A070909	Oversensing	08/01/2024	
									A090801	Signal Artifact/Noise	08/01/2024	
14728	2024	Death	47265	Dual-chamber implantable pacemaker, rate-responsive	CT1685	Cardiac pacemakers	E011903	Syncope/Fainting	A071102	Failure to Interrogate	12/01/2024	12/01/2024
									A071204	No Pacing	12/01/2024	12/01/2024
							E0601	Arrhythmia	A071102	Failure to Interrogate	12/01/2024	12/01/2024
									A071204	No Pacing	12/01/2024	12/01/2024
							E0742	Respiratory Failure	A071102	Failure to Interrogate	12/01/2024	12/01/2024
									A071204	No Pacing	12/01/2024	12/01/2024
							E2402	Appropriate Clinical Signs, Symptoms and Conditions Term/Code not Available	A071102	Failure to Interrogate	12/01/2024	12/01/2024
									A071204	No Pacing	12/01/2024	12/01/2024
							F02	Death	A071102	Failure to Interrogate	12/01/2024	12/01/2024
									A071204	No Pacing	12/01/2024	12/01/2024
					CT257	Implantable cardiac pacemakers	E011903	Syncope/Fainting	A071102	Failure to Interrogate	12/01/2024	12/01/2024
									A071204	No Pacing	12/01/2024	12/01/2024
							E0601	Arrhythmia	A071102	Failure to Interrogate	12/01/2024	12/01/2024
									A071204	No Pacing	12/01/2024	12/01/2024
							E0742	Respiratory Failure	A071102	Failure to Interrogate	12/01/2024	12/01/2024
									A071204	No Pacing	12/01/2024	12/01/2024
							E2402	Appropriate Clinical Signs, Symptoms and Conditions Term/Code not Available	A071102	Failure to Interrogate	12/01/2024	12/01/2024
									A071204	No Pacing	12/01/2024	12/01/2024
							F02	Death	A071102	Failure to Interrogate	12/01/2024	12/01/2024
									A071204	No Pacing	12/01/2024	12/01/2024
					CT2866	Implantable cardiac electrical stimulators	E011903	Syncope/Fainting	A071102	Failure to Interrogate	12/01/2024	12/01/2024
									A071204	No Pacing	12/01/2024	12/01/2024
							E0601	Arrhythmia	A071102	Failure to Interrogate	12/01/2024	12/01/2024
									A071204	No Pacing	12/01/2024	12/01/2024
							E0742	Respiratory Failure	A071102	Failure to Interrogate	12/01/2024	12/01/2024
									A071204	No Pacing	12/01/2024	12/01/2024
							E2402	Appropriate Clinical Signs, Symptoms and Conditions Term/Code not Available	A071102	Failure to Interrogate	12/01/2024	12/01/2024
									A071204	No Pacing	12/01/2024	12/01/2024
							F02	Death	A071102	Failure to Interrogate	12/01/2024	12/01/2024
									A071204	No Pacing	12/01/2024	12/01/2024
14729	2024	All other reportable incidents	47265	Dual-chamber implantable pacemaker, rate-responsive	CT1685	Cardiac pacemakers	E060110	Ventricular Fibrillation	A24	No Apparent Device Problem	08/01/2024	08/01/2024
							F08	Hospitalization or Prolonged Hospitalization	A24	No Apparent Device Problem	08/01/2024	08/01/2024
							F1203	Life Threatening Illness or Injury	A24	No Apparent Device Problem	08/01/2024	08/01/2024
							F2306	Resuscitation	A24	No Apparent Device Problem	08/01/2024	08/01/2024
					CT257	Implantable cardiac pacemakers	E060110	Ventricular Fibrillation	A24	No Apparent Device Problem	08/01/2024	08/01/2024
							F08	Hospitalization or Prolonged Hospitalization	A24	No Apparent Device Problem	08/01/2024	08/01/2024
							F1203	Life Threatening Illness or Injury	A24	No Apparent Device Problem	08/01/2024	08/01/2024
							F2306	Resuscitation	A24	No Apparent Device Problem	08/01/2024	08/01/2024
					CT2866	Implantable cardiac electrical stimulators	E060110	Ventricular Fibrillation	A24	No Apparent Device Problem	08/01/2024	08/01/2024
							F08	Hospitalization or Prolonged Hospitalization	A24	No Apparent Device Problem	08/01/2024	08/01/2024
							F1203	Life Threatening Illness or Injury	A24	No Apparent Device Problem	08/01/2024	08/01/2024
							F2306	Resuscitation	A24	No Apparent Device Problem	08/01/2024	08/01/2024
14730	2024	All other reportable incidents	35852	Single-chamber implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
									A0908	Incorrect, Inadequate or Imprecise Result or Readings		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
									A0908	Incorrect, Inadequate or Imprecise Result or Readings		
14731	2024		35852	Single-chamber implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A0202	Defective Component		
									A07	Electrical /Electronic Property Problem		
							F26	No Health Consequences or Impact	A0202	Defective Component		
									A07	Electrical /Electronic Property Problem		
14732	2024	All other reportable incidents	47265	Dual-chamber implantable pacemaker, rate-responsive	CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem		
							F1901	Additional Surgery	A24	No Apparent Device Problem		
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem		
							F1901	Additional Surgery	A24	No Apparent Device Problem		
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem		
							F1901	Additional Surgery	A24	No Apparent Device Problem		
14733	2024	All other reportable incidents	17882	Physiologic-monitoring defibrillation system	CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070803	Failure to Power Up		
							F27	Problem Identified During Non-Clinical Procedure	A070803	Failure to Power Up		
14734	2024	All other reportable incidents	47265	Dual-chamber implantable pacemaker, rate-responsive	CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	22/12/2023	22/12/2023
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	22/12/2023	22/12/2023
							F1903	Device Explantation	A070504	Premature Discharge of Battery	22/12/2023	22/12/2023
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	22/12/2023	22/12/2023
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	22/12/2023	22/12/2023
							F1903	Device Explantation	A070504	Premature Discharge of Battery	22/12/2023	22/12/2023
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	22/12/2023	22/12/2023
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	22/12/2023	22/12/2023
							F1903	Device Explantation	A070504	Premature Discharge of Battery	22/12/2023	22/12/2023
14735	2024	All other reportable incidents	47265	Dual-chamber implantable pacemaker, rate-responsive	CT1685	Cardiac pacemakers	E0601	Arrhythmia	A071102	Failure to Interrogate	15/01/2024	15/01/2024
									A071204	No Pacing	15/01/2024	15/01/2024
							F1901	Additional Surgery	A071102	Failure to Interrogate	15/01/2024	15/01/2024
									A071204	No Pacing	15/01/2024	15/01/2024
					CT257	Implantable cardiac pacemakers	E0601	Arrhythmia	A071102	Failure to Interrogate	15/01/2024	15/01/2024
									A071204	No Pacing	15/01/2024	15/01/2024
							F1901	Additional Surgery	A071102	Failure to Interrogate	15/01/2024	15/01/2024
									A071204	No Pacing	15/01/2024	15/01/2024
					CT2866	Implantable cardiac electrical stimulators	E0601	Arrhythmia	A071102	Failure to Interrogate	15/01/2024	15/01/2024

									A071204	No Pacing	15/01/2024	15/01/2024
							F1901	Additional Surgery	A071102	Failure to Interrogate	15/01/2024	15/01/2024
									A071204	No Pacing	15/01/2024	15/01/2024
14736	2024	All other reportable incidents	37265	Dual-chamber implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	07/02/2024	07/02/2024
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	07/02/2024	07/02/2024
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	07/02/2024	07/02/2024
14737	2024	All other reportable incidents	35852	Single-chamber implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E210401	Shock from Patient Lead(s)	A070909	Oversensing		
									A071302	Inappropriate/Inadequate Shock/Stimulation		
									A1301	Failure to Read Input Signal		
									A23	Use of Device Problem		
							F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A070909	Oversensing		
									A071302	Inappropriate/Inadequate Shock/Stimulation		
									A1301	Failure to Read Input Signal		
									A23	Use of Device Problem		
14738	2024	All other reportable incidents	47265	Dual-chamber implantable pacemaker, rate-responsive	CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A070101	Failure to Capture		
							F26	No Health Consequences or Impact	A070101	Failure to Capture		
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A070101	Failure to Capture		
							F26	No Health Consequences or Impact	A070101	Failure to Capture		
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A070101	Failure to Capture		
							F26	No Health Consequences or Impact	A070101	Failure to Capture		
14739	2024	All other reportable incidents	35852	Single-chamber implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A072201	High Impedance		
							F26	No Health Consequences or Impact	A072201	High Impedance		
14740	2024	All other reportable incidents	37265	Dual-chamber implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A0709	Device Sensing Problem		
							F26	No Health Consequences or Impact	A0709	Device Sensing Problem		
14741	2024	All other reportable incidents	47265	Dual-chamber implantable pacemaker, rate-responsive	CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	14/12/2023	14/12/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	14/12/2023	14/12/2023
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	14/12/2023	14/12/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	14/12/2023	14/12/2023
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	14/12/2023	14/12/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	14/12/2023	14/12/2023
14742	2024	All other reportable incidents	35852	Single-chamber implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
									A0908	Incorrect, Inadequate or Imprecise Result or Readings		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
									A0908	Incorrect, Inadequate or Imprecise Result or Readings		
14743	2024	All other reportable incidents	47263	Cardiac resynchronization therapy implantable pacemaker	CT1685	Cardiac pacemakers	E060101	Asystole	A070909	Oversensing		
									A0712	Pacing Problem		
									A072201	High Impedance		
									A090801	Signal Artifact/Noise		
							F1001	Absence of Treatment	A070909	Oversensing		
									A0712	Pacing Problem		
									A072201	High Impedance		
									A090801	Signal Artifact/Noise		
					CT257	Implantable cardiac pacemakers	E060101	Asystole	A070909	Oversensing		
									A0712	Pacing Problem		
									A072201	High Impedance		
									A090801	Signal Artifact/Noise		
							F1001	Absence of Treatment	A070909	Oversensing		
									A0712	Pacing Problem		
									A072201	High Impedance		
									A090801	Signal Artifact/Noise		
					CT2866	Implantable cardiac electrical stimulators	E060101	Asystole	A070909	Oversensing		
									A0712	Pacing Problem		
									A072201	High Impedance		
									A090801	Signal Artifact/Noise		
							F1001	Absence of Treatment	A070909	Oversensing		
									A0712	Pacing Problem		
									A072201	High Impedance		
									A090801	Signal Artifact/Noise		
14744	2024	All other reportable incidents	47265	Dual-chamber implantable pacemaker, rate-responsive	CT1685	Cardiac pacemakers	E011903	Syncope/Fainting	A27	Appropriate Device Problem Term/Code not Available		
							F1203	Life Threatening Illness or Injury	A27	Appropriate Device Problem Term/Code not Available		
							F23	Unexpected Medical Intervention	A27	Appropriate Device Problem Term/Code not Available		
					CT257	Implantable cardiac pacemakers	E011903	Syncope/Fainting	A27	Appropriate Device Problem Term/Code not Available		
							F1203	Life Threatening Illness or Injury	A27	Appropriate Device Problem Term/Code not Available		
							F23	Unexpected Medical Intervention	A27	Appropriate Device Problem Term/Code not Available		
					CT2866	Implantable cardiac electrical stimulators	E011903	Syncope/Fainting	A27	Appropriate Device Problem Term/Code not Available		
							F1203	Life Threatening Illness or Injury	A27	Appropriate Device Problem Term/Code not Available		
							F23	Unexpected Medical Intervention	A27	Appropriate Device Problem Term/Code not Available		
14745	2024	Death	17882	Physiologic-monitoring defibrillation system	CT1684	Cardiac pacemakers and associated devices	E0623	Unspecified Heart Problem	A071301	Failure to Deliver Shock/Stimulation		
							F02	Death	A071301	Failure to Deliver Shock/Stimulation		
14746	2024	Death	35852	Single-chamber implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E060109	Tachycardia	A0401	Break		
									A0713	Defibrillation/Stimulation Problem		
									A072201	High Impedance		
									A072202	Low Impedance		
									A09	Output Problem		
									A090805	Incorrect Measurement		
							E060110	Ventricular Fibrillation	A0401	Break		
									A0713	Defibrillation/Stimulation Problem		
									A072201	High Impedance		
									A072202	Low Impedance		

									A09	Output Problem			
									A090805	Incorrect Measurement			
							F02	Death	A0401	Break			
									A0713	Defibrillation/Stimulation Problem			
									A072201	High Impedance			
									A072202	Low Impedance			
									A09	Output Problem			
									A090805	Incorrect Measurement			
							F0801	Intensive Care	A0401	Break			
									A0713	Defibrillation/Stimulation Problem			
									A072201	High Impedance			
									A072202	Low Impedance			
									A09	Output Problem			
									A090805	Incorrect Measurement			
							F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A0401	Break			
									A0713	Defibrillation/Stimulation Problem			
									A072201	High Impedance			
									A072202	Low Impedance			
									A09	Output Problem			
									A090805	Incorrect Measurement			
							F23	Unexpected Medical Intervention	A0401	Break			
									A0713	Defibrillation/Stimulation Problem			
									A072201	High Impedance			
									A072202	Low Impedance			
									A09	Output Problem			
									A090805	Incorrect Measurement			
							F2306	Resuscitation	A0401	Break			
									A0713	Defibrillation/Stimulation Problem			
									A072201	High Impedance			
									A072202	Low Impedance			
									A09	Output Problem			
									A090805	Incorrect Measurement			
14747	2024	All other reportable incidents	37265	Dual-chamber implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	01/12/2023	31/12/2023	
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	01/12/2023	31/12/2023	
							F12	Serious Injury/ Illness/ Impairment	A070504	Premature Discharge of Battery	01/12/2023	31/12/2023	
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	01/12/2023	31/12/2023	
14748	2024	Death	35852	Single-chamber implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E210401	Shock from Patient Lead(s)	A0709	Device Sensing Problem			
									A071302	Inappropriate/Inadequate Shock/Stimulation			
									A090811	Unable to Obtain Readings			
							F02	Death	A0709	Device Sensing Problem			
									A071302	Inappropriate/Inadequate Shock/Stimulation			
									A090811	Unable to Obtain Readings			
							F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A0709	Device Sensing Problem			
									A071302	Inappropriate/Inadequate Shock/Stimulation			
									A090811	Unable to Obtain Readings			
							F1203	Life Threatening Illness or Injury	A0709	Device Sensing Problem			
									A071302	Inappropriate/Inadequate Shock/Stimulation			
									A090811	Unable to Obtain Readings			
							F23	Unexpected Medical Intervention	A0709	Device Sensing Problem			
									A071302	Inappropriate/Inadequate Shock/Stimulation			
									A090811	Unable to Obtain Readings			
14749	2024	All other reportable incidents	47205	Cardiac pulse generator programmer	CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A150105	Key or Button Unresponsive/not Working			
							F26	No Health Consequences or Impact	A150105	Key or Button Unresponsive/not Working			
14750	2024	All other reportable incidents	47263	Cardiac resynchronization therapy implantable pacemaker	CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A0908	Incorrect, Inadequate or Imprecise Result or Readings	17/01/2024	17/01/2024	
							F08	Hospitalization or Prolonged Hospitalization	A0908	Incorrect, Inadequate or Imprecise Result or Readings	17/01/2024	17/01/2024	
							F1905	Device Revision or Replacement	A0908	Incorrect, Inadequate or Imprecise Result or Readings	17/01/2024	17/01/2024	
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A0908	Incorrect, Inadequate or Imprecise Result or Readings	17/01/2024	17/01/2024	
							F08	Hospitalization or Prolonged Hospitalization	A0908	Incorrect, Inadequate or Imprecise Result or Readings	17/01/2024	17/01/2024	
							F1905	Device Revision or Replacement	A0908	Incorrect, Inadequate or Imprecise Result or Readings	17/01/2024	17/01/2024	
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A0908	Incorrect, Inadequate or Imprecise Result or Readings	17/01/2024	17/01/2024	
							F08	Hospitalization or Prolonged Hospitalization	A0908	Incorrect, Inadequate or Imprecise Result or Readings	17/01/2024	17/01/2024	
							F1905	Device Revision or Replacement	A0908	Incorrect, Inadequate or Imprecise Result or Readings	17/01/2024	17/01/2024	
14751	2024	All other reportable incidents	37265	Dual-chamber implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E2104	Electric Shock	A070910	Undersensing			
									A071302	Inappropriate/Inadequate Shock/Stimulation			
							F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A070910	Undersensing			
									A071302	Inappropriate/Inadequate Shock/Stimulation			
14752	2024	All other reportable incidents	47265	Dual-chamber implantable pacemaker, rate-responsive	CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery			
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery			
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery			
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery			
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery			
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery			
14753	2024	All other reportable incidents	17882	Physiologic-monitoring defibrillation system	CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A07	Electrical /Electronic Property Problem			
							F27	Problem Identified During Non-Clinical Procedure	A07	Electrical /Electronic Property Problem			
14754	2024	All other reportable incidents	47265	Dual-chamber implantable pacemaker, rate-responsive	CT1685	Cardiac pacemakers	E0601	Arrhythmia	A0712	Pacing Problem			
							E060109	Tachycardia	A0712	Pacing Problem			
							F12	Serious Injury/ Illness/ Impairment	A0712	Pacing Problem			
							F23	Unexpected Medical Intervention	A0712	Pacing Problem			
					CT257	Implantable cardiac pacemakers	E0601	Arrhythmia	A0712	Pacing Problem			
							E060109	Tachycardia	A0712	Pacing Problem			
							F12	Serious Injury/ Illness/ Impairment	A0712	Pacing Problem			

					CT2866	Implantable cardiac electrical stimulators	F23	Unexpected Medical Intervention	A0712	Pacing Problem			
							E0601	Arrhythmia	A0712	Pacing Problem			
							E060109	Tachycardia	A0712	Pacing Problem			
							F12	Serious Injury/ Illness/ Impairment	A0712	Pacing Problem			
							F23	Unexpected Medical Intervention	A0712	Pacing Problem			
14755	2024	All other reportable incidents	47265	Dual-chamber implantable pacemaker, rate-responsive	CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A0908	Incorrect, Inadequate or Imprecise Result or Readings			
							F26	No Health Consequences or Impact	A0908	Incorrect, Inadequate or Imprecise Result or Readings			
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A0908	Incorrect, Inadequate or Imprecise Result or Readings			
							F26	No Health Consequences or Impact	A0908	Incorrect, Inadequate or Imprecise Result or Readings			
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A0908	Incorrect, Inadequate or Imprecise Result or Readings			
							F26	No Health Consequences or Impact	A0908	Incorrect, Inadequate or Imprecise Result or Readings			
14756	2024	All other reportable incidents	47263	Cardiac resynchronization therapy implantable pacemaker	CT1685	Cardiac pacemakers	E0112	Dizziness	A0908	Incorrect, Inadequate or Imprecise Result or Readings			
							E0138	Undesired Nerve Stimulation	A0908	Incorrect, Inadequate or Imprecise Result or Readings			
							F26	No Health Consequences or Impact	A0908	Incorrect, Inadequate or Imprecise Result or Readings			
					CT257	Implantable cardiac pacemakers	E0112	Dizziness	A0908	Incorrect, Inadequate or Imprecise Result or Readings			
							E0138	Undesired Nerve Stimulation	A0908	Incorrect, Inadequate or Imprecise Result or Readings			
							F26	No Health Consequences or Impact	A0908	Incorrect, Inadequate or Imprecise Result or Readings			
					CT2866	Implantable cardiac electrical stimulators	E0112	Dizziness	A0908	Incorrect, Inadequate or Imprecise Result or Readings			
							E0138	Undesired Nerve Stimulation	A0908	Incorrect, Inadequate or Imprecise Result or Readings			
							F26	No Health Consequences or Impact	A0908	Incorrect, Inadequate or Imprecise Result or Readings			
14757	2024	All other reportable incidents	47265	Dual-chamber implantable pacemaker, rate-responsive	CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	12/01/2024	12/01/2024	
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	12/01/2024	12/01/2024	
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	12/01/2024	12/01/2024	
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	12/01/2024	12/01/2024	
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	12/01/2024	12/01/2024	
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	12/01/2024	12/01/2024	
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	12/01/2024	12/01/2024	
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	12/01/2024	12/01/2024	
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	12/01/2024	12/01/2024	
14758	2024	All other reportable incidents	37265	Dual-chamber implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem			
							F26	No Health Consequences or Impact	A0705	Battery Problem			
14759	2024	All other reportable incidents	17882	Physiologic-monitoring defibrillation system	CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A07	Electrical /Electronic Property Problem			
							F27	Problem Identified During Non-Clinical Procedure	A07	Electrical /Electronic Property Problem			
14760	2024	All other reportable incidents	35852	Single-chamber implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem	01/11/2023	01/11/2023	
									A070504	Premature Discharge of Battery	01/11/2023	01/11/2023	
									A0908	Incorrect, Inadequate or Imprecise Result or Readings	01/11/2023	01/11/2023	
							F08	Hospitalization or Prolonged Hospitalization	A0705	Battery Problem	01/11/2023	01/11/2023	
									A070504	Premature Discharge of Battery	01/11/2023	01/11/2023	
									A0908	Incorrect, Inadequate or Imprecise Result or Readings	01/11/2023	01/11/2023	
							F1905	Device Revision or Replacement	A0705	Battery Problem	01/11/2023	01/11/2023	
									A070504	Premature Discharge of Battery	01/11/2023	01/11/2023	
									A0908	Incorrect, Inadequate or Imprecise Result or Readings	01/11/2023	01/11/2023	
14761	2024	All other reportable incidents	47265	Dual-chamber implantable pacemaker, rate-responsive	CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A0701	Capturing Problem			
							F26	No Health Consequences or Impact	A0701	Capturing Problem			
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A0701	Capturing Problem			
							F26	No Health Consequences or Impact	A0701	Capturing Problem			
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A0701	Capturing Problem			
							F26	No Health Consequences or Impact	A0701	Capturing Problem			
14762	2024	All other reportable incidents	47205	Cardiac pulse generator programmer	CT1684	Cardiac pacemakers and associated devices	E011903	Syncope/Fainting	A1104	Computer Operating System Problem			
							F12	Serious Injury/ Illness/ Impairment	A1104	Computer Operating System Problem			
							F23	Unexpected Medical Intervention	A1104	Computer Operating System Problem			
14763	2024	All other reportable incidents	60789	Intracardiac pacemaker	CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A0504	Leak/Splash			
							F26	No Health Consequences or Impact	A0504	Leak/Splash			
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A0504	Leak/Splash			
							F26	No Health Consequences or Impact	A0504	Leak/Splash			
					CT2818	Catheterization/Cannulation/Fluid path devices	E2403	No Clinical Signs, Symptoms or Conditions	A0504	Leak/Splash			
							F26	No Health Consequences or Impact	A0504	Leak/Splash			
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A0504	Leak/Splash			
							F26	No Health Consequences or Impact	A0504	Leak/Splash			
14764	2024	All other reportable incidents	47263	Cardiac resynchronization therapy implantable pacemaker	CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A0908	Incorrect, Inadequate or Imprecise Result or Readings	13/10/2023	13/10/2023	
							F08	Hospitalization or Prolonged Hospitalization	A0908	Incorrect, Inadequate or Imprecise Result or Readings	13/10/2023	13/10/2023	
							F1905	Device Revision or Replacement	A0908	Incorrect, Inadequate or Imprecise Result or Readings	13/10/2023	13/10/2023	
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A0908	Incorrect, Inadequate or Imprecise Result or Readings	13/10/2023	13/10/2023	
							F08	Hospitalization or Prolonged Hospitalization	A0908	Incorrect, Inadequate or Imprecise Result or Readings	13/10/2023	13/10/2023	
							F1905	Device Revision or Replacement	A0908	Incorrect, Inadequate or Imprecise Result or Readings	13/10/2023	13/10/2023	
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A0908	Incorrect, Inadequate or Imprecise Result or Readings	13/10/2023	13/10/2023	
							F08	Hospitalization or Prolonged Hospitalization	A0908	Incorrect, Inadequate or Imprecise Result or Readings	13/10/2023	13/10/2023	

							F1905	Device Revision or Replacement	A0908	Incorrect, Inadequate or Imprecise Result or Readings	13/10/2023	13/10/2023
14765	2024	All other reportable incidents	35852	Single-chamber implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
14766	2024	All other reportable incidents	63188	Left-ventricular wireless pacing system sensor/transmitter	CT141	Electrodes/Leads and associated devices			A0708	Power Problem		
					CT145	Cardiovascular devices			A0708	Power Problem		
					CT1577	Wireless monitoring system transmitters			A0708	Power Problem		
					CT1684	Cardiac pacemakers and associated devices			A0708	Power Problem		
					CT2317	Transmitters/Receivers/Transceivers			A0708	Power Problem		
					CT233	Surgical			A0708	Power Problem		
					CT2406	Implantable			A0708	Power Problem		
					CT2407	Active implantable			A0708	Power Problem		
					CT255	Transducers and associated devices			A0708	Power Problem		
					CT2726	Physiological monitoring/measuring devices			A0708	Power Problem		
					CT2727	Active implant devices/ancillaries			A0708	Power Problem		
					CT2750	Implantable cardiac electrical stimulators and associated devices			A0708	Power Problem		
					CT2885	Physiological monitoring/measuring system transmitters/receivers/transceivers			A0708	Power Problem		
					CT2985	Active			A0708	Power Problem		
					CT3001	Implantable metal			A0708	Power Problem		
					CT3002	Implantable synthetic polymer			A0708	Power Problem		
					CT3006	Implantable inorganic material			A0708	Power Problem		
					CT314	Electrically powered			A0708	Power Problem		
					CT321	Long-term surgical invasive			A0708	Power Problem		
					CT384	Sensors and associated devices			A0708	Power Problem		
					CT449	Generators and associated devices			A0708	Power Problem		
					CT981	Single use			A0708	Power Problem		
					CT983	Surgical invasive			A0708	Power Problem		
					CT994	Sensors			A0708	Power Problem		
14767	2024	All other reportable incidents	47265	Dual-chamber implantable pacemaker, rate-responsive	CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	15/11/2023	15/11/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	15/11/2023	15/11/2023
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	15/11/2023	15/11/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	15/11/2023	15/11/2023
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	15/11/2023	15/11/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	15/11/2023	15/11/2023
14768	2024	All other reportable incidents	47265	Dual-chamber implantable pacemaker, rate-responsive	CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem	22/01/2024	22/01/2024
									A0908	Incorrect, Inadequate or Imprecise Result or Readings	22/01/2024	22/01/2024
							F08	Hospitalization or Prolonged Hospitalization	A0705	Battery Problem	22/01/2024	22/01/2024
									A0908	Incorrect, Inadequate or Imprecise Result or Readings	22/01/2024	22/01/2024
							F1905	Device Revision or Replacement	A0705	Battery Problem	22/01/2024	22/01/2024
									A0908	Incorrect, Inadequate or Imprecise Result or Readings	22/01/2024	22/01/2024
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem	22/01/2024	22/01/2024
									A0908	Incorrect, Inadequate or Imprecise Result or Readings	22/01/2024	22/01/2024
							F08	Hospitalization or Prolonged Hospitalization	A0705	Battery Problem	22/01/2024	22/01/2024
									A0908	Incorrect, Inadequate or Imprecise Result or Readings	22/01/2024	22/01/2024
							F1905	Device Revision or Replacement	A0705	Battery Problem	22/01/2024	22/01/2024
									A0908	Incorrect, Inadequate or Imprecise Result or Readings	22/01/2024	22/01/2024
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem	22/01/2024	22/01/2024
									A0908	Incorrect, Inadequate or Imprecise Result or Readings	22/01/2024	22/01/2024
							F08	Hospitalization or Prolonged Hospitalization	A0705	Battery Problem	22/01/2024	22/01/2024
									A0908	Incorrect, Inadequate or Imprecise Result or Readings	22/01/2024	22/01/2024
							F1905	Device Revision or Replacement	A0705	Battery Problem	22/01/2024	22/01/2024
									A0908	Incorrect, Inadequate or Imprecise Result or Readings	22/01/2024	22/01/2024
14769	2024	All other reportable incidents	17882	Physiologic-monitoring defibrillation system	CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A0908	Incorrect, Inadequate or Imprecise Result or Readings		
							F26	No Health Consequences or Impact	A0908	Incorrect, Inadequate or Imprecise Result or Readings		
14770	2024	All other reportable incidents	47205	Cardiac pulse generator programmer	CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A09	Output Problem		
									A1104	Computer Operating System Problem		
									A13	Communication or Transmission Problem		
							F26	No Health Consequences or Impact	A09	Output Problem		
									A1104	Computer Operating System Problem		
									A13	Communication or Transmission Problem		
14771	2024	All other reportable incidents	35852	Single-chamber implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E2104	Electric Shock	A0709	Device Sensing Problem	23/04/2024	23/04/2024
									A070909	Oversensing	23/04/2024	23/04/2024
									A071302	Inappropriate/Inadequate Shock/Stimulation	23/04/2024	23/04/2024
							F08	Hospitalization or Prolonged Hospitalization	A0709	Device Sensing Problem	23/04/2024	23/04/2024
									A070909	Oversensing	23/04/2024	23/04/2024
									A071302	Inappropriate/Inadequate Shock/Stimulation	23/04/2024	23/04/2024
							F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A0709	Device Sensing Problem	23/04/2024	23/04/2024
									A070909	Oversensing	23/04/2024	23/04/2024
									A071302	Inappropriate/Inadequate Shock/Stimulation	23/04/2024	23/04/2024
							F1903	Device Explantation	A0709	Device Sensing Problem	23/04/2024	23/04/2024
									A070909	Oversensing	23/04/2024	23/04/2024
									A071302	Inappropriate/Inadequate Shock/Stimulation	23/04/2024	23/04/2024
							F2203	Imaging Required	A0709	Device Sensing Problem	23/04/2024	23/04/2024
									A070909	Oversensing	23/04/2024	23/04/2024
									A071302	Inappropriate/Inadequate Shock/Stimulation	23/04/2024	23/04/2024
14772	2024		47265	Dual-chamber implantable pacemaker, rate-responsive	CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem		
							F1901	Additional Surgery	A24	No Apparent Device Problem		
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem		

							F1901	Additional Surgery	A24	No Apparent Device Problem		
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem		
							F1901	Additional Surgery	A24	No Apparent Device Problem		
14773	2024	All other reportable incidents	17882	Physiologic-monitoring defibrillation system	CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A07	Electrical /Electronic Property Problem		
							F27	Problem Identified During Non-Clinical Procedure	A07	Electrical /Electronic Property Problem		
14774	2024	All other reportable incidents	47265	Dual-chamber implantable pacemaker, rate-responsive	CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
14775	2024		47265	Dual-chamber implantable pacemaker, rate-responsive	CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	05/12/2023	05/12/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	05/12/2023	05/12/2023
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	05/12/2023	05/12/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	05/12/2023	05/12/2023
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	05/12/2023	05/12/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	05/12/2023	05/12/2023
14776	2024	All other reportable incidents	17882	Physiologic-monitoring defibrillation system	CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A0908	Incorrect, Inadequate or Imprecise Result or Readings		
							F27	Problem Identified During Non-Clinical Procedure	A0908	Incorrect, Inadequate or Imprecise Result or Readings		
14777	2024	All other reportable incidents	47270	Cardiac resynchronization therapy implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	25/01/2024	25/01/2024
							F1901	Additional Surgery	A070504	Premature Discharge of Battery	25/01/2024	25/01/2024
14778	2024	All other reportable incidents	47270	Cardiac resynchronization therapy implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	23/01/2024	23/01/2024
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	23/01/2024	23/01/2024
							F12	Serious Injury/ Illness/ Impairment	A070504	Premature Discharge of Battery	23/01/2024	23/01/2024
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	23/01/2024	23/01/2024
14779	2024	All other reportable incidents	47265	Dual-chamber implantable pacemaker, rate-responsive	CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A0908	Incorrect, Inadequate or Imprecise Result or Readings		
							F08	Hospitalization or Prolonged Hospitalization	A0908	Incorrect, Inadequate or Imprecise Result or Readings		
							F1905	Device Revision or Replacement	A0908	Incorrect, Inadequate or Imprecise Result or Readings		
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A0908	Incorrect, Inadequate or Imprecise Result or Readings		
							F08	Hospitalization or Prolonged Hospitalization	A0908	Incorrect, Inadequate or Imprecise Result or Readings		
							F1905	Device Revision or Replacement	A0908	Incorrect, Inadequate or Imprecise Result or Readings		
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A0908	Incorrect, Inadequate or Imprecise Result or Readings		
							F08	Hospitalization or Prolonged Hospitalization	A0908	Incorrect, Inadequate or Imprecise Result or Readings		
							F1905	Device Revision or Replacement	A0908	Incorrect, Inadequate or Imprecise Result or Readings		
14780	2024	All other reportable incidents	47265	Dual-chamber implantable pacemaker, rate-responsive	CT1685	Cardiac pacemakers	E0112	Dizziness	A070101	Failure to Capture	22/01/2024	22/01/2024
									A0908	Incorrect, Inadequate or Imprecise Result or Readings	22/01/2024	22/01/2024
							E060101	Asystole	A070101	Failure to Capture	22/01/2024	22/01/2024
									A0908	Incorrect, Inadequate or Imprecise Result or Readings	22/01/2024	22/01/2024
							F08	Hospitalization or Prolonged Hospitalization	A070101	Failure to Capture	22/01/2024	22/01/2024
									A0908	Incorrect, Inadequate or Imprecise Result or Readings	22/01/2024	22/01/2024
							F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A070101	Failure to Capture	22/01/2024	22/01/2024
									A0908	Incorrect, Inadequate or Imprecise Result or Readings	22/01/2024	22/01/2024
							F1905	Device Revision or Replacement	A070101	Failure to Capture	22/01/2024	22/01/2024
									A0908	Incorrect, Inadequate or Imprecise Result or Readings	22/01/2024	22/01/2024
					CT257	Implantable cardiac pacemakers	E0112	Dizziness	A070101	Failure to Capture	22/01/2024	22/01/2024
									A0908	Incorrect, Inadequate or Imprecise Result or Readings	22/01/2024	22/01/2024
							E060101	Asystole	A070101	Failure to Capture	22/01/2024	22/01/2024
									A0908	Incorrect, Inadequate or Imprecise Result or Readings	22/01/2024	22/01/2024
							F08	Hospitalization or Prolonged Hospitalization	A070101	Failure to Capture	22/01/2024	22/01/2024
									A0908	Incorrect, Inadequate or Imprecise Result or Readings	22/01/2024	22/01/2024
							F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A070101	Failure to Capture	22/01/2024	22/01/2024
									A0908	Incorrect, Inadequate or Imprecise Result or Readings	22/01/2024	22/01/2024
							F1905	Device Revision or Replacement	A070101	Failure to Capture	22/01/2024	22/01/2024
									A0908	Incorrect, Inadequate or Imprecise Result or Readings	22/01/2024	22/01/2024
					CT2866	Implantable cardiac electrical stimulators	E0112	Dizziness	A070101	Failure to Capture	22/01/2024	22/01/2024
									A0908	Incorrect, Inadequate or Imprecise Result or Readings	22/01/2024	22/01/2024
							E060101	Asystole	A070101	Failure to Capture	22/01/2024	22/01/2024
									A0908	Incorrect, Inadequate or Imprecise Result or Readings	22/01/2024	22/01/2024
							F08	Hospitalization or Prolonged Hospitalization	A070101	Failure to Capture	22/01/2024	22/01/2024
									A0908	Incorrect, Inadequate or Imprecise Result or Readings	22/01/2024	22/01/2024
							F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A070101	Failure to Capture	22/01/2024	22/01/2024
									A0908	Incorrect, Inadequate or Imprecise Result or Readings	22/01/2024	22/01/2024
							F1905	Device Revision or Replacement	A070101	Failure to Capture	22/01/2024	22/01/2024

									A0908	Incorrect, Inadequate or Imprecise Result or Readings	22/01/2024	22/01/2024
14781	2024		47265	Dual-chamber implantable pacemaker, rate-responsive	CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	09/01/2024	09/01/2024
							F1901	Additional Surgery	A24	No Apparent Device Problem	09/01/2024	09/01/2024
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	09/01/2024	09/01/2024
							F1901	Additional Surgery	A24	No Apparent Device Problem	09/01/2024	09/01/2024
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	09/01/2024	09/01/2024
							F1901	Additional Surgery	A24	No Apparent Device Problem	09/01/2024	09/01/2024
14782	2024		47265	Dual-chamber implantable pacemaker, rate-responsive	CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	23/01/2024	23/01/2024
							F1901	Additional Surgery	A24	No Apparent Device Problem	23/01/2024	23/01/2024
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	23/01/2024	23/01/2024
							F1901	Additional Surgery	A24	No Apparent Device Problem	23/01/2024	23/01/2024
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	23/01/2024	23/01/2024
							F1901	Additional Surgery	A24	No Apparent Device Problem	23/01/2024	23/01/2024
14783	2024	All other reportable incidents	17882	Physiologic-monitoring defibrillation system	CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070912	Incorrect Interpretation of Signal		
							F27	Problem Identified During Non-Clinical Procedure	A070912	Incorrect Interpretation of Signal		
14784	2024	All other reportable incidents	47265	Dual-chamber implantable pacemaker, rate-responsive	CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A071102	Failure to Interrogate	26/01/2024	26/01/2024
									A071204	No Pacing	26/01/2024	26/01/2024
							F1901	Additional Surgery	A071102	Failure to Interrogate	26/01/2024	26/01/2024
									A071204	No Pacing	26/01/2024	26/01/2024
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A071102	Failure to Interrogate	26/01/2024	26/01/2024
									A071204	No Pacing	26/01/2024	26/01/2024
							F1901	Additional Surgery	A071102	Failure to Interrogate	26/01/2024	26/01/2024
									A071204	No Pacing	26/01/2024	26/01/2024
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A071102	Failure to Interrogate	26/01/2024	26/01/2024
									A071204	No Pacing	26/01/2024	26/01/2024
							F1901	Additional Surgery	A071102	Failure to Interrogate	26/01/2024	26/01/2024
									A071204	No Pacing	26/01/2024	26/01/2024
14785	2024	All other reportable incidents	37265	Dual-chamber implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A070909	Oversensing		
									A090801	Signal Artifact/Noise		
									A1301	Failure to Read Input Signal		
							F26	No Health Consequences or Impact	A070909	Oversensing		
									A090801	Signal Artifact/Noise		
									A1301	Failure to Read Input Signal		
14786	2024	All other reportable incidents	47265	Dual-chamber implantable pacemaker, rate-responsive	CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A0908	Incorrect, Inadequate or Imprecise Result or Readings	26/01/2024	26/01/2024
							F08	Hospitalization or Prolonged Hospitalization	A0908	Incorrect, Inadequate or Imprecise Result or Readings	26/01/2024	26/01/2024
							F1905	Device Revision or Replacement	A0908	Incorrect, Inadequate or Imprecise Result or Readings	26/01/2024	26/01/2024
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A0908	Incorrect, Inadequate or Imprecise Result or Readings	26/01/2024	26/01/2024
							F08	Hospitalization or Prolonged Hospitalization	A0908	Incorrect, Inadequate or Imprecise Result or Readings	26/01/2024	26/01/2024
							F1905	Device Revision or Replacement	A0908	Incorrect, Inadequate or Imprecise Result or Readings	26/01/2024	26/01/2024
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A0908	Incorrect, Inadequate or Imprecise Result or Readings	26/01/2024	26/01/2024
							F08	Hospitalization or Prolonged Hospitalization	A0908	Incorrect, Inadequate or Imprecise Result or Readings	26/01/2024	26/01/2024
							F1905	Device Revision or Replacement	A0908	Incorrect, Inadequate or Imprecise Result or Readings	26/01/2024	26/01/2024
14787	2024	All other reportable incidents	47267	Single-chamber implantable pacemaker, rate-responsive	CT1685	Cardiac pacemakers	E0717	Dyspnea	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
							E2312	Fatigue	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
							F24	Insufficient Health Impact Information	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
					CT257	Implantable cardiac pacemakers	E0717	Dyspnea	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
							E2312	Fatigue	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
							F24	Insufficient Health Impact Information	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
					CT2866	Implantable cardiac electrical stimulators	E0717	Dyspnea	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
							E2312	Fatigue	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
							F24	Insufficient Health Impact Information	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
14788	2024	All other reportable incidents	17882	Physiologic-monitoring defibrillation system	CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A07	Electrical /Electronic Property Problem		
							F27	Problem Identified During Non-Clinical Procedure	A07	Electrical /Electronic Property Problem		
14789	2024		47265	Dual-chamber implantable pacemaker, rate-responsive	CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	24/11/2023	24/11/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	24/11/2023	24/11/2023
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	24/11/2023	24/11/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	24/11/2023	24/11/2023
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	24/11/2023	24/11/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	24/11/2023	24/11/2023
14790	2024		47265	Dual-chamber implantable pacemaker, rate-responsive	CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	24/11/2023	24/11/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	24/11/2023	24/11/2023
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	24/11/2023	24/11/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	24/11/2023	24/11/2023
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	24/11/2023	24/11/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	24/11/2023	24/11/2023
14791	2024		35852	Single-chamber implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem		
							F1901	Additional Surgery	A0705	Battery Problem		
							F2301	Additional Device Required	A0705	Battery Problem		
14792	2024	All other reportable incidents	37265	Dual-chamber implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A09	Output Problem		
									A090811	Unable to Obtain Readings		
							F26	No Health Consequences or Impact	A09	Output Problem		

								A090811	Unable to Obtain Readings		
14793	2024	All other reportable incidents	47265	Dual-chamber implantable pacemaker, rate-responsive	CT1685	Cardiac pacemakers	E0112	Dizziness	A070909	Oversensing	27/01/2024	27/01/2024
									A0712	Pacing Problem	27/01/2024	27/01/2024
									A0908	Incorrect, Inadequate or Imprecise Result or Readings	27/01/2024	27/01/2024
							E011902	Presyncope	A070909	Oversensing	27/01/2024	27/01/2024
									A0712	Pacing Problem	27/01/2024	27/01/2024
									A0908	Incorrect, Inadequate or Imprecise Result or Readings	27/01/2024	27/01/2024
							E060104	Bradycardia	A070909	Oversensing	27/01/2024	27/01/2024
									A0712	Pacing Problem	27/01/2024	27/01/2024
									A0908	Incorrect, Inadequate or Imprecise Result or Readings	27/01/2024	27/01/2024
							F08	Hospitalization or Prolonged Hospitalization	A070909	Oversensing	27/01/2024	27/01/2024
									A0712	Pacing Problem	27/01/2024	27/01/2024
									A0908	Incorrect, Inadequate or Imprecise Result or Readings	27/01/2024	27/01/2024
							F1001	Absence of Treatment	A070909	Oversensing	27/01/2024	27/01/2024
									A0712	Pacing Problem	27/01/2024	27/01/2024
									A0908	Incorrect, Inadequate or Imprecise Result or Readings	27/01/2024	27/01/2024
							F1905	Device Revision or Replacement	A070909	Oversensing	27/01/2024	27/01/2024
									A0712	Pacing Problem	27/01/2024	27/01/2024
									A0908	Incorrect, Inadequate or Imprecise Result or Readings	27/01/2024	27/01/2024
					CT257	Implantable cardiac pacemakers	E0112	Dizziness	A070909	Oversensing	27/01/2024	27/01/2024
									A0712	Pacing Problem	27/01/2024	27/01/2024
									A0908	Incorrect, Inadequate or Imprecise Result or Readings	27/01/2024	27/01/2024
							E011902	Presyncope	A070909	Oversensing	27/01/2024	27/01/2024
									A0712	Pacing Problem	27/01/2024	27/01/2024
									A0908	Incorrect, Inadequate or Imprecise Result or Readings	27/01/2024	27/01/2024
							E060104	Bradycardia	A070909	Oversensing	27/01/2024	27/01/2024
									A0712	Pacing Problem	27/01/2024	27/01/2024
									A0908	Incorrect, Inadequate or Imprecise Result or Readings	27/01/2024	27/01/2024
							F08	Hospitalization or Prolonged Hospitalization	A070909	Oversensing	27/01/2024	27/01/2024
									A0712	Pacing Problem	27/01/2024	27/01/2024
									A0908	Incorrect, Inadequate or Imprecise Result or Readings	27/01/2024	27/01/2024
							F1001	Absence of Treatment	A070909	Oversensing	27/01/2024	27/01/2024
									A0712	Pacing Problem	27/01/2024	27/01/2024
									A0908	Incorrect, Inadequate or Imprecise Result or Readings	27/01/2024	27/01/2024
							F1905	Device Revision or Replacement	A070909	Oversensing	27/01/2024	27/01/2024
									A0712	Pacing Problem	27/01/2024	27/01/2024
									A0908	Incorrect, Inadequate or Imprecise Result or Readings	27/01/2024	27/01/2024
					CT2866	Implantable cardiac electrical stimulators	E0112	Dizziness	A070909	Oversensing	27/01/2024	27/01/2024
									A0712	Pacing Problem	27/01/2024	27/01/2024
									A0908	Incorrect, Inadequate or Imprecise Result or Readings	27/01/2024	27/01/2024
							E011902	Presyncope	A070909	Oversensing	27/01/2024	27/01/2024
									A0712	Pacing Problem	27/01/2024	27/01/2024
									A0908	Incorrect, Inadequate or Imprecise Result or Readings	27/01/2024	27/01/2024
							E060104	Bradycardia	A070909	Oversensing	27/01/2024	27/01/2024
									A0712	Pacing Problem	27/01/2024	27/01/2024
									A0908	Incorrect, Inadequate or Imprecise Result or Readings	27/01/2024	27/01/2024
							F08	Hospitalization or Prolonged Hospitalization	A070909	Oversensing	27/01/2024	27/01/2024
									A0712	Pacing Problem	27/01/2024	27/01/2024
									A0908	Incorrect, Inadequate or Imprecise Result or Readings	27/01/2024	27/01/2024
							F1001	Absence of Treatment	A070909	Oversensing	27/01/2024	27/01/2024
									A0712	Pacing Problem	27/01/2024	27/01/2024
									A0908	Incorrect, Inadequate or Imprecise Result or Readings	27/01/2024	27/01/2024
							F1905	Device Revision or Replacement	A070909	Oversensing	27/01/2024	27/01/2024
									A0712	Pacing Problem	27/01/2024	27/01/2024
									A0908	Incorrect, Inadequate or Imprecise Result or Readings	27/01/2024	27/01/2024
14794	2024	All other reportable incidents	47263	Cardiac resynchronization therapy implantable pacemaker	CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A0908	Incorrect, Inadequate or Imprecise Result or Readings	26/11/2023	26/11/2023
							F08	Hospitalization or Prolonged Hospitalization	A0908	Incorrect, Inadequate or Imprecise Result or Readings	26/11/2023	26/11/2023
							F1905	Device Revision or Replacement	A0908	Incorrect, Inadequate or Imprecise Result or Readings	26/11/2023	26/11/2023
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A0908	Incorrect, Inadequate or Imprecise Result or Readings	26/11/2023	26/11/2023
							F08	Hospitalization or Prolonged Hospitalization	A0908	Incorrect, Inadequate or Imprecise Result or Readings	26/11/2023	26/11/2023
							F1905	Device Revision or Replacement	A0908	Incorrect, Inadequate or Imprecise Result or Readings	26/11/2023	26/11/2023
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A0908	Incorrect, Inadequate or Imprecise Result or Readings	26/11/2023	26/11/2023
							F08	Hospitalization or Prolonged Hospitalization	A0908	Incorrect, Inadequate or Imprecise Result or Readings	26/11/2023	26/11/2023
							F1905	Device Revision or Replacement	A0908	Incorrect, Inadequate or Imprecise Result or Readings	26/11/2023	26/11/2023
14795	2024		47265	Dual-chamber implantable pacemaker, rate-responsive	CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	29/01/2024	29/01/2024
							F1901	Additional Surgery	A24	No Apparent Device Problem	29/01/2024	29/01/2024
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	29/01/2024	29/01/2024

							F1901	Additional Surgery	A24	No Apparent Device Problem	29/01/2024	29/01/2024
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	29/01/2024	29/01/2024
							F1901	Additional Surgery	A24	No Apparent Device Problem	29/01/2024	29/01/2024
14796	2024		47265	Dual-chamber implantable pacemaker, rate-responsive	CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A0203	Defective Device		
									A071205	Pacemaker Found in Back-Up Mode		
							F1903	Device Explantation	A0203	Defective Device		
									A071205	Pacemaker Found in Back-Up Mode		
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A0203	Defective Device		
									A071205	Pacemaker Found in Back-Up Mode		
							F1903	Device Explantation	A0203	Defective Device		
									A071205	Pacemaker Found in Back-Up Mode		
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A0203	Defective Device		
									A071205	Pacemaker Found in Back-Up Mode		
							F1903	Device Explantation	A0203	Defective Device		
									A071205	Pacemaker Found in Back-Up Mode		
14797	2024	All other reportable incidents	47267	Single-chamber implantable pacemaker, rate-responsive	CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
14798	2024	All other reportable incidents	17882	Physiologic-monitoring defibrillation system	CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F27	Problem Identified During Non-Clinical Procedure	A070504	Premature Discharge of Battery		
14799	2024	All other reportable incidents	35852	Single-chamber implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
14800	2024	All other reportable incidents	47265	Dual-chamber implantable pacemaker, rate-responsive	CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A0908	Incorrect, Inadequate or Imprecise Result or Readings	23/01/2024	23/01/2024
									A1107	Data Problem	23/01/2024	23/01/2024
							F08	Hospitalization or Prolonged Hospitalization	A0908	Incorrect, Inadequate or Imprecise Result or Readings	23/01/2024	23/01/2024
									A1107	Data Problem	23/01/2024	23/01/2024
							F1905	Device Revision or Replacement	A0908	Incorrect, Inadequate or Imprecise Result or Readings	23/01/2024	23/01/2024
									A1107	Data Problem	23/01/2024	23/01/2024
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A0908	Incorrect, Inadequate or Imprecise Result or Readings	23/01/2024	23/01/2024
									A1107	Data Problem	23/01/2024	23/01/2024
							F08	Hospitalization or Prolonged Hospitalization	A0908	Incorrect, Inadequate or Imprecise Result or Readings	23/01/2024	23/01/2024
									A1107	Data Problem	23/01/2024	23/01/2024
							F1905	Device Revision or Replacement	A0908	Incorrect, Inadequate or Imprecise Result or Readings	23/01/2024	23/01/2024
									A1107	Data Problem	23/01/2024	23/01/2024
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A0908	Incorrect, Inadequate or Imprecise Result or Readings	23/01/2024	23/01/2024
									A1107	Data Problem	23/01/2024	23/01/2024
							F08	Hospitalization or Prolonged Hospitalization	A0908	Incorrect, Inadequate or Imprecise Result or Readings	23/01/2024	23/01/2024
									A1107	Data Problem	23/01/2024	23/01/2024
							F1905	Device Revision or Replacement	A0908	Incorrect, Inadequate or Imprecise Result or Readings	23/01/2024	23/01/2024
									A1107	Data Problem	23/01/2024	23/01/2024
14801	2024	All other reportable incidents	47265	Dual-chamber implantable pacemaker, rate-responsive	CT1685	Cardiac pacemakers	E0112	Dizziness	A0723	Unexpected Mode Switch	06/02/2023	06/02/2023
							E0601	Arrhythmia	A0723	Unexpected Mode Switch	06/02/2023	06/02/2023
							F1901	Additional Surgery	A0723	Unexpected Mode Switch	06/02/2023	06/02/2023
					CT257	Implantable cardiac pacemakers	E0112	Dizziness	A0723	Unexpected Mode Switch	06/02/2023	06/02/2023
							E0601	Arrhythmia	A0723	Unexpected Mode Switch	06/02/2023	06/02/2023
							F1901	Additional Surgery	A0723	Unexpected Mode Switch	06/02/2023	06/02/2023
					CT2866	Implantable cardiac electrical stimulators	E0112	Dizziness	A0723	Unexpected Mode Switch	06/02/2023	06/02/2023
							E0601	Arrhythmia	A0723	Unexpected Mode Switch	06/02/2023	06/02/2023
							F1901	Additional Surgery	A0723	Unexpected Mode Switch	06/02/2023	06/02/2023
14802	2024	All other reportable incidents	47265	Dual-chamber implantable pacemaker, rate-responsive	CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A0203	Defective Device	22/12/2023	22/12/2023
									A071102	Failure to Interrogate	22/12/2023	22/12/2023
							F1901	Additional Surgery	A0203	Defective Device	22/12/2023	22/12/2023
									A071102	Failure to Interrogate	22/12/2023	22/12/2023
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A0203	Defective Device	22/12/2023	22/12/2023
									A071102	Failure to Interrogate	22/12/2023	22/12/2023
							F1901	Additional Surgery	A0203	Defective Device	22/12/2023	22/12/2023
									A071102	Failure to Interrogate	22/12/2023	22/12/2023
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A0203	Defective Device	22/12/2023	22/12/2023
									A071102	Failure to Interrogate	22/12/2023	22/12/2023
							F1901	Additional Surgery	A0203	Defective Device	22/12/2023	22/12/2023
									A071102	Failure to Interrogate	22/12/2023	22/12/2023
14803	2024		17882	Physiologic-monitoring defibrillation system	CT1684	Cardiac pacemakers and associated devices			A071301	Failure to Deliver Shock/Stimulation		
14804	2024		47267	Single-chamber implantable pacemaker, rate-responsive	CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	28/11/2023	28/11/2023
							F1901	Additional Surgery	A070504	Premature Discharge of Battery	28/11/2023	28/11/2023
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	28/11/2023	28/11/2023
							F1901	Additional Surgery	A070504	Premature Discharge of Battery	28/11/2023	28/11/2023
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	28/11/2023	28/11/2023
							F1901	Additional Surgery	A070504	Premature Discharge of Battery	28/11/2023	28/11/2023
14805	2024	All other reportable incidents	47265	Dual-chamber implantable pacemaker, rate-responsive	CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		

					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
14806	2024		47265	Dual-chamber implantable pacemaker, rate-responsive	CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	16/01/2024	16/01/2024
							F1901	Additional Surgery	A24	No Apparent Device Problem	16/01/2024	16/01/2024
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	16/01/2024	16/01/2024
							F1901	Additional Surgery	A24	No Apparent Device Problem	16/01/2024	16/01/2024
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	16/01/2024	16/01/2024
							F1901	Additional Surgery	A24	No Apparent Device Problem	16/01/2024	16/01/2024
14807	2024	All other reportable incidents	35852	Single-chamber implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E210401	Shock from Patient Lead(s)	A070909	Oversensing		
									A071302	Inappropriate/Inadequate Shock/Stimulation		
									A090801	Signal Artifact/Noise		
									A1301	Failure to Read Input Signal		
							F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A070909	Oversensing		
									A071302	Inappropriate/Inadequate Shock/Stimulation		
									A090801	Signal Artifact/Noise		
									A1301	Failure to Read Input Signal		
14808	2024	All other reportable incidents	37265	Dual-chamber implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A0508	Noise, Audible		
									A072201	High Impedance		
							F26	No Health Consequences or Impact	A0508	Noise, Audible		
									A072201	High Impedance		
14809	2024	All other reportable incidents	47265	Dual-chamber implantable pacemaker, rate-responsive	CT1685	Cardiac pacemakers	E2103	Device Overstimulation of Tissue	A0908	Incorrect, Inadequate or Imprecise Result or Readings	15/01/2024	15/01/2024
							F08	Hospitalization or Prolonged Hospitalization	A0908	Incorrect, Inadequate or Imprecise Result or Readings	15/01/2024	15/01/2024
							F1905	Device Revision or Replacement	A0908	Incorrect, Inadequate or Imprecise Result or Readings	15/01/2024	15/01/2024
					CT257	Implantable cardiac pacemakers	E2103	Device Overstimulation of Tissue	A0908	Incorrect, Inadequate or Imprecise Result or Readings	15/01/2024	15/01/2024
							F08	Hospitalization or Prolonged Hospitalization	A0908	Incorrect, Inadequate or Imprecise Result or Readings	15/01/2024	15/01/2024
							F1905	Device Revision or Replacement	A0908	Incorrect, Inadequate or Imprecise Result or Readings	15/01/2024	15/01/2024
					CT2866	Implantable cardiac electrical stimulators	E2103	Device Overstimulation of Tissue	A0908	Incorrect, Inadequate or Imprecise Result or Readings	15/01/2024	15/01/2024
							F08	Hospitalization or Prolonged Hospitalization	A0908	Incorrect, Inadequate or Imprecise Result or Readings	15/01/2024	15/01/2024
							F1905	Device Revision or Replacement	A0908	Incorrect, Inadequate or Imprecise Result or Readings	15/01/2024	15/01/2024
14810	2024	All other reportable incidents	35852	Single-chamber implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E210401	Shock from Patient Lead(s)	A0709	Device Sensing Problem		
									A070909	Oversensing		
									A071302	Inappropriate/Inadequate Shock/Stimulation		
							F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A0709	Device Sensing Problem		
									A070909	Oversensing		
									A071302	Inappropriate/Inadequate Shock/Stimulation		
							F15	Recognized Device or Procedural Complication	A0709	Device Sensing Problem		
									A070909	Oversensing		
									A071302	Inappropriate/Inadequate Shock/Stimulation		
14811	2024		47265	Dual-chamber implantable pacemaker, rate-responsive	CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	19/01/2024	19/01/2024
							F1901	Additional Surgery	A24	No Apparent Device Problem	19/01/2024	19/01/2024
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	19/01/2024	19/01/2024
							F1901	Additional Surgery	A24	No Apparent Device Problem	19/01/2024	19/01/2024
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	19/01/2024	19/01/2024
							F1901	Additional Surgery	A24	No Apparent Device Problem	19/01/2024	19/01/2024
14812	2024		47265	Dual-chamber implantable pacemaker, rate-responsive	CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	23/01/2024	23/01/2024
							F1901	Additional Surgery	A24	No Apparent Device Problem	23/01/2024	23/01/2024
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	23/01/2024	23/01/2024
							F1901	Additional Surgery	A24	No Apparent Device Problem	23/01/2024	23/01/2024
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	23/01/2024	23/01/2024
							F1901	Additional Surgery	A24	No Apparent Device Problem	23/01/2024	23/01/2024
14813	2024	All other reportable incidents	17882	Physiologic-monitoring defibrillation system	CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F27	Problem Identified During Non-Clinical Procedure	A070504	Premature Discharge of Battery		
14814	2024	All other reportable incidents	35852	Single-chamber implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	22/02/2024	22/02/2024
									A0908	Incorrect, Inadequate or Imprecise Result or Readings	22/02/2024	22/02/2024
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	22/02/2024	22/02/2024
									A0908	Incorrect, Inadequate or Imprecise Result or Readings	22/02/2024	22/02/2024
							F1903	Device Explantation	A070504	Premature Discharge of Battery	22/02/2024	22/02/2024
									A0908	Incorrect, Inadequate or Imprecise Result or Readings	22/02/2024	22/02/2024
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	22/02/2024	22/02/2024
									A0908	Incorrect, Inadequate or Imprecise Result or Readings	22/02/2024	22/02/2024
14815	2024		47265	Dual-chamber implantable pacemaker, rate-responsive	CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem		
							F1901	Additional Surgery	A24	No Apparent Device Problem		
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem		
							F1901	Additional Surgery	A24	No Apparent Device Problem		
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem		
							F1901	Additional Surgery	A24	No Apparent Device Problem		
14816	2024	All other reportable incidents	47265	Dual-chamber implantable pacemaker, rate-responsive	CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
14817	2024		47270	Cardiac resynchronization therapy implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A070101	Failure to Capture		
							F23	Unexpected Medical Intervention	A070101	Failure to Capture		
14818	2024	All other reportable incidents	47265	Dual-chamber implantable pacemaker, rate-responsive	CT141	Electrodes/Leads and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070909	Oversensing		
									A090801	Signal Artifact/Noise		

							F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A070909	Oversensing		
									A090801	Signal Artifact/Noise		
					CT145	Cardiovascular devices	E2403	No Clinical Signs, Symptoms or Conditions	A070909	Oversensing		
									A090801	Signal Artifact/Noise		
							F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A070909	Oversensing		
									A090801	Signal Artifact/Noise		
					CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070909	Oversensing		
									A090801	Signal Artifact/Noise		
							F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A070909	Oversensing		
									A090801	Signal Artifact/Noise		
					CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A070909	Oversensing		
									A090801	Signal Artifact/Noise		
							F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A070909	Oversensing		
									A090801	Signal Artifact/Noise		
					CT233	Surgical	E2403	No Clinical Signs, Symptoms or Conditions	A070909	Oversensing		
									A090801	Signal Artifact/Noise		
							F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A070909	Oversensing		
									A090801	Signal Artifact/Noise		
					CT2406	Implantable	E2403	No Clinical Signs, Symptoms or Conditions	A070909	Oversensing		
									A090801	Signal Artifact/Noise		
							F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A070909	Oversensing		
									A090801	Signal Artifact/Noise		
					CT2407	Active implantable	E2403	No Clinical Signs, Symptoms or Conditions	A070909	Oversensing		
									A090801	Signal Artifact/Noise		
							F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A070909	Oversensing		
									A090801	Signal Artifact/Noise		
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A070909	Oversensing		
									A090801	Signal Artifact/Noise		
							F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A070909	Oversensing		
									A090801	Signal Artifact/Noise		
					CT2727	Active implant devices/ancillaries	E2403	No Clinical Signs, Symptoms or Conditions	A070909	Oversensing		
									A090801	Signal Artifact/Noise		
							F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A070909	Oversensing		
									A090801	Signal Artifact/Noise		
					CT2750	Implantable cardiac electrical stimulators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070909	Oversensing		
									A090801	Signal Artifact/Noise		
							F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A070909	Oversensing		
									A090801	Signal Artifact/Noise		
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A070909	Oversensing		
									A090801	Signal Artifact/Noise		
							F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A070909	Oversensing		
									A090801	Signal Artifact/Noise		
					CT2985	Active	E2403	No Clinical Signs, Symptoms or Conditions	A070909	Oversensing		
									A090801	Signal Artifact/Noise		
							F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A070909	Oversensing		
									A090801	Signal Artifact/Noise		
					CT3001	Implantable metal	E2403	No Clinical Signs, Symptoms or Conditions	A070909	Oversensing		
									A090801	Signal Artifact/Noise		
							F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A070909	Oversensing		
									A090801	Signal Artifact/Noise		
					CT3002	Implantable synthetic polymer	E2403	No Clinical Signs, Symptoms or Conditions	A070909	Oversensing		
									A090801	Signal Artifact/Noise		
							F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A070909	Oversensing		
									A090801	Signal Artifact/Noise		
					CT3006	Implantable inorganic material	E2403	No Clinical Signs, Symptoms or Conditions	A070909	Oversensing		
									A090801	Signal Artifact/Noise		
							F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A070909	Oversensing		
									A090801	Signal Artifact/Noise		
					CT314	Electrically powered	E2403	No Clinical Signs, Symptoms or Conditions	A070909	Oversensing		
									A090801	Signal Artifact/Noise		
							F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A070909	Oversensing		
									A090801	Signal Artifact/Noise		
					CT321	Long-term surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A070909	Oversensing		
									A090801	Signal Artifact/Noise		
							F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A070909	Oversensing		
									A090801	Signal Artifact/Noise		
					CT449	Generators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070909	Oversensing		
									A090801	Signal Artifact/Noise		
							F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A070909	Oversensing		
									A090801	Signal Artifact/Noise		
					CT981	Single use	E2403	No Clinical Signs, Symptoms or Conditions	A070909	Oversensing		
									A090801	Signal Artifact/Noise		
							F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A070909	Oversensing		
									A090801	Signal Artifact/Noise		
					CT983	Surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A070909	Oversensing		
									A090801	Signal Artifact/Noise		
							F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A070909	Oversensing		
									A090801	Signal Artifact/Noise		
14819	2024	All other reportable incidents			CT1684	Cardiac pacemakers and associated devices	E210401	Shock from Patient Lead(s)	A0709	Device Sensing Problem		
									A070909	Oversensing		
									A071302	Inappropriate/Inadequate Shock/Stimulation		
							F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A0709	Device Sensing Problem		
									A070909	Oversensing		
									A071302	Inappropriate/Inadequate Shock/Stimulation		
							F2203	Imaging Required	A0709	Device Sensing Problem		
									A070909	Oversensing		
									A071302	Inappropriate/Inadequate Shock/Stimulation		
14820	2024	All other reportable incidents	35852	Single-chamber implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	12/10/2023	12/10/2023

							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	12/10/2023	12/10/2023
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	12/10/2023	12/10/2023
14821	2024	All other reportable incidents	17882	Physiologic-monitoring defibrillation system	CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A071301	Failure to Deliver Shock/Stimulation		
							F27	Problem Identified During Non-Clinical Procedure	A071301	Failure to Deliver Shock/Stimulation		
14822	2024	All other reportable incidents	47265	Dual-chamber implantable pacemaker, rate-responsive	CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A0908	Incorrect, Inadequate or Imprecise Result or Readings	21/11/2023	21/11/2023
							F08	Hospitalization or Prolonged Hospitalization	A0908	Incorrect, Inadequate or Imprecise Result or Readings	21/11/2023	21/11/2023
							F1905	Device Revision or Replacement	A0908	Incorrect, Inadequate or Imprecise Result or Readings	21/11/2023	21/11/2023
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A0908	Incorrect, Inadequate or Imprecise Result or Readings	21/11/2023	21/11/2023
							F08	Hospitalization or Prolonged Hospitalization	A0908	Incorrect, Inadequate or Imprecise Result or Readings	21/11/2023	21/11/2023
							F1905	Device Revision or Replacement	A0908	Incorrect, Inadequate or Imprecise Result or Readings	21/11/2023	21/11/2023
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A0908	Incorrect, Inadequate or Imprecise Result or Readings	21/11/2023	21/11/2023
							F08	Hospitalization or Prolonged Hospitalization	A0908	Incorrect, Inadequate or Imprecise Result or Readings	21/11/2023	21/11/2023
							F1905	Device Revision or Replacement	A0908	Incorrect, Inadequate or Imprecise Result or Readings	21/11/2023	21/11/2023
14823	2024	All other reportable incidents	47270	Cardiac resynchronization therapy implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A071102	Failure to Interrogate	10/11/2023	10/11/2023
							F08	Hospitalization or Prolonged Hospitalization	A071102	Failure to Interrogate	10/11/2023	10/11/2023
							F12	Serious Injury/ Illness/ Impairment	A071102	Failure to Interrogate	10/11/2023	10/11/2023
							F1905	Device Revision or Replacement	A071102	Failure to Interrogate	10/11/2023	10/11/2023
							F26	No Health Consequences or Impact	A071102	Failure to Interrogate	10/11/2023	10/11/2023
14824	2024		47265	Dual-chamber implantable pacemaker, rate-responsive	CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	02/02/2024	02/02/2024
							F1901	Additional Surgery	A24	No Apparent Device Problem	02/02/2024	02/02/2024
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	02/02/2024	02/02/2024
							F1901	Additional Surgery	A24	No Apparent Device Problem	02/02/2024	02/02/2024
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	02/02/2024	02/02/2024
							F1901	Additional Surgery	A24	No Apparent Device Problem	02/02/2024	02/02/2024
14825	2024		47270	Cardiac resynchronization therapy implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem		
							F1903	Device Explantation	A0705	Battery Problem		
14826	2024	All other reportable incidents	47267	Single-chamber implantable pacemaker, rate-responsive	CT1685	Cardiac pacemakers	E2103	Device Overstimulation of Tissue	A0908	Incorrect, Inadequate or Imprecise Result or Readings		
							F08	Hospitalization or Prolonged Hospitalization	A0908	Incorrect, Inadequate or Imprecise Result or Readings		
							F15	Recognized Device or Procedural Complication	A0908	Incorrect, Inadequate or Imprecise Result or Readings		
					CT257	Implantable cardiac pacemakers	E2103	Device Overstimulation of Tissue	A0908	Incorrect, Inadequate or Imprecise Result or Readings		
							F08	Hospitalization or Prolonged Hospitalization	A0908	Incorrect, Inadequate or Imprecise Result or Readings		
							F15	Recognized Device or Procedural Complication	A0908	Incorrect, Inadequate or Imprecise Result or Readings		
					CT2866	Implantable cardiac electrical stimulators	E2103	Device Overstimulation of Tissue	A0908	Incorrect, Inadequate or Imprecise Result or Readings		
							F08	Hospitalization or Prolonged Hospitalization	A0908	Incorrect, Inadequate or Imprecise Result or Readings		
							F15	Recognized Device or Procedural Complication	A0908	Incorrect, Inadequate or Imprecise Result or Readings		
14827	2024	All other reportable incidents	47265	Dual-chamber implantable pacemaker, rate-responsive	CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	22/01/2024	22/01/2024
							F1901	Additional Surgery	A24	No Apparent Device Problem	22/01/2024	22/01/2024
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	22/01/2024	22/01/2024
							F1901	Additional Surgery	A24	No Apparent Device Problem	22/01/2024	22/01/2024
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	22/01/2024	22/01/2024
							F1901	Additional Surgery	A24	No Apparent Device Problem	22/01/2024	22/01/2024
14828	2024	All other reportable incidents	47265	Dual-chamber implantable pacemaker, rate-responsive	CT1685	Cardiac pacemakers	E0112	Dizziness	A070101	Failure to Capture		
									A072201	High Impedance		
							F15	Recognized Device or Procedural Complication	A070101	Failure to Capture		
									A072201	High Impedance		
					CT257	Implantable cardiac pacemakers	E0112	Dizziness	A070101	Failure to Capture		
									A072201	High Impedance		
							F15	Recognized Device or Procedural Complication	A070101	Failure to Capture		
									A072201	High Impedance		
					CT2866	Implantable cardiac electrical stimulators	E0112	Dizziness	A070101	Failure to Capture		
									A072201	High Impedance		
							F15	Recognized Device or Procedural Complication	A070101	Failure to Capture		
									A072201	High Impedance		
14829	2024	All other reportable incidents	35852	Single-chamber implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem	08/03/2024	08/03/2024
									A070504	Premature Discharge of Battery	08/03/2024	08/03/2024
							F08	Hospitalization or Prolonged Hospitalization	A0705	Battery Problem	08/03/2024	08/03/2024
									A070504	Premature Discharge of Battery	08/03/2024	08/03/2024
							F1905	Device Revision or Replacement	A0705	Battery Problem	08/03/2024	08/03/2024
									A070504	Premature Discharge of Battery	08/03/2024	08/03/2024
14830	2024	All other reportable incidents	17882	Physiologic-monitoring defibrillation system	CT1684	Cardiac pacemakers and associated devices			A071301	Failure to Deliver Shock/Stimulation		
14831	2024	All other reportable incidents	47263	Cardiac resynchronization therapy implantable pacemaker	CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A070909	Oversensing		
									A070910	Undersensing		
							F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A070909	Oversensing		
									A070910	Undersensing		
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A070909	Oversensing		
									A070910	Undersensing		
							F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A070909	Oversensing		
									A070910	Undersensing		
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A070909	Oversensing		
									A070910	Undersensing		
							F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A070909	Oversensing		
									A070910	Undersensing		
14832	2024	All other reportable incidents	35852	Single-chamber implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A0908	Incorrect, Inadequate or Imprecise Result or Readings		

									A1107	Data Problem		
									A13	Communication or Transmission Problem		
							F26	No Health Consequences or Impact	A0908	Incorrect, Inadequate or Imprecise Result or Readings		
									A1107	Data Problem		
									A13	Communication or Transmission Problem		
14833	2024	All other reportable incidents	47265	Dual-chamber implantable pacemaker, rate-responsive	CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A0723	Unexpected Mode Switch		
							F1901	Additional Surgery	A0723	Unexpected Mode Switch		
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A0723	Unexpected Mode Switch		
							F1901	Additional Surgery	A0723	Unexpected Mode Switch		
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A0723	Unexpected Mode Switch		
							F1901	Additional Surgery	A0723	Unexpected Mode Switch		
14834	2024		47265	Dual-chamber implantable pacemaker, rate-responsive	CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	12/01/2024	12/01/2024
							F1901	Additional Surgery	A24	No Apparent Device Problem	12/01/2024	12/01/2024
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	12/01/2024	12/01/2024
							F1901	Additional Surgery	A24	No Apparent Device Problem	12/01/2024	12/01/2024
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	12/01/2024	12/01/2024
							F1901	Additional Surgery	A24	No Apparent Device Problem	12/01/2024	12/01/2024
14835	2024	All other reportable incidents	17882	Physiologic-monitoring defibrillation system	CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A1603	Failure of Device to Self-Test		
							F27	Problem Identified During Non-Clinical Procedure	A1603	Failure of Device to Self-Test		
14836	2024	All other reportable incidents	37265	Dual-chamber implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A070909	Oversensing		
							F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A070909	Oversensing		
14837	2024	All other reportable incidents	17882	Physiologic-monitoring defibrillation system	CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A071301	Failure to Deliver Shock/Stimulation		
							F27	Problem Identified During Non-Clinical Procedure	A071301	Failure to Deliver Shock/Stimulation		
14838	2024	All other reportable incidents	60789	Intracardiac pacemaker	CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A0701	Capturing Problem		
									A070104	Unstable Capture Threshold		
									A0722	Impedance Problem		
							F12	Serious Injury/ Illness/ Impairment	A0701	Capturing Problem		
									A070104	Unstable Capture Threshold		
									A0722	Impedance Problem		
							F23	Unexpected Medical Intervention	A0701	Capturing Problem		
									A070104	Unstable Capture Threshold		
									A0722	Impedance Problem		
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A0701	Capturing Problem		
									A070104	Unstable Capture Threshold		
									A0722	Impedance Problem		
							F12	Serious Injury/ Illness/ Impairment	A0701	Capturing Problem		
									A070104	Unstable Capture Threshold		
									A0722	Impedance Problem		
							F23	Unexpected Medical Intervention	A0701	Capturing Problem		
									A070104	Unstable Capture Threshold		
									A0722	Impedance Problem		
					CT2818	Catheterization/Cannulation/Fluid path devices	E2403	No Clinical Signs, Symptoms or Conditions	A0701	Capturing Problem		
									A070104	Unstable Capture Threshold		
									A0722	Impedance Problem		
							F12	Serious Injury/ Illness/ Impairment	A0701	Capturing Problem		
									A070104	Unstable Capture Threshold		
									A0722	Impedance Problem		
							F23	Unexpected Medical Intervention	A0701	Capturing Problem		
									A070104	Unstable Capture Threshold		
									A0722	Impedance Problem		
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A0701	Capturing Problem		
									A070104	Unstable Capture Threshold		
									A0722	Impedance Problem		
							F12	Serious Injury/ Illness/ Impairment	A0701	Capturing Problem		
									A070104	Unstable Capture Threshold		
									A0722	Impedance Problem		
							F23	Unexpected Medical Intervention	A0701	Capturing Problem		
									A070104	Unstable Capture Threshold		
									A0722	Impedance Problem		
14839	2024	All other reportable incidents	47265	Dual-chamber implantable pacemaker, rate-responsive	CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A070909	Oversensing		
									A072201	High Impedance		
							F26	No Health Consequences or Impact	A070909	Oversensing		
									A072201	High Impedance		
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A070909	Oversensing		
									A072201	High Impedance		
							F26	No Health Consequences or Impact	A070909	Oversensing		
									A072201	High Impedance		
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A070909	Oversensing		
									A072201	High Impedance		
							F26	No Health Consequences or Impact	A070909	Oversensing		
									A072201	High Impedance		
14840	2024	All other reportable incidents	47465	Cardiac contractility modulation system	CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A150105	Key or Button Unresponsive/not Working		
							F26	No Health Consequences or Impact	A150105	Key or Button Unresponsive/not Working		
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A150105	Key or Button Unresponsive/not Working		
							F26	No Health Consequences or Impact	A150105	Key or Button Unresponsive/not Working		
14841	2024	All other reportable incidents	47263	Cardiac resynchronization therapy implantable pacemaker	CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A070909	Oversensing		
									A090801	Signal Artifact/Noise		
							F2203	Imaging Required	A070909	Oversensing		
									A090801	Signal Artifact/Noise		
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A070909	Oversensing		
									A090801	Signal Artifact/Noise		
							F2203	Imaging Required	A070909	Oversensing		
									A090801	Signal Artifact/Noise		
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A070909	Oversensing		
									A090801	Signal Artifact/Noise		
							F2203	Imaging Required	A070909	Oversensing		
									A090801	Signal Artifact/Noise		

14842	2024		47265	Dual-chamber implantable pacemaker, rate-responsive	CT141	Electrodes/Leads and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	08/02/2024	08/02/2024
							F1901	Additional Surgery	A24	No Apparent Device Problem	08/02/2024	08/02/2024
					CT145	Cardiovascular devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	08/02/2024	08/02/2024
							F1901	Additional Surgery	A24	No Apparent Device Problem	08/02/2024	08/02/2024
					CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	08/02/2024	08/02/2024
							F1901	Additional Surgery	A24	No Apparent Device Problem	08/02/2024	08/02/2024
					CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	08/02/2024	08/02/2024
							F1901	Additional Surgery	A24	No Apparent Device Problem	08/02/2024	08/02/2024
					CT233	Surgical	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	08/02/2024	08/02/2024
							F1901	Additional Surgery	A24	No Apparent Device Problem	08/02/2024	08/02/2024
					CT2406	Implantable	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	08/02/2024	08/02/2024
							F1901	Additional Surgery	A24	No Apparent Device Problem	08/02/2024	08/02/2024
					CT2407	Active implantable	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	08/02/2024	08/02/2024
							F1901	Additional Surgery	A24	No Apparent Device Problem	08/02/2024	08/02/2024
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	08/02/2024	08/02/2024
							F1901	Additional Surgery	A24	No Apparent Device Problem	08/02/2024	08/02/2024
					CT2727	Active implant devices/ancillaries	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	08/02/2024	08/02/2024
							F1901	Additional Surgery	A24	No Apparent Device Problem	08/02/2024	08/02/2024
					CT2750	Implantable cardiac electrical stimulators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	08/02/2024	08/02/2024
							F1901	Additional Surgery	A24	No Apparent Device Problem	08/02/2024	08/02/2024
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	08/02/2024	08/02/2024
							F1901	Additional Surgery	A24	No Apparent Device Problem	08/02/2024	08/02/2024
					CT2985	Active	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	08/02/2024	08/02/2024
							F1901	Additional Surgery	A24	No Apparent Device Problem	08/02/2024	08/02/2024
					CT3001	Implantable metal	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	08/02/2024	08/02/2024
							F1901	Additional Surgery	A24	No Apparent Device Problem	08/02/2024	08/02/2024
					CT3002	Implantable synthetic polymer	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	08/02/2024	08/02/2024
							F1901	Additional Surgery	A24	No Apparent Device Problem	08/02/2024	08/02/2024
					CT3006	Implantable inorganic material	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	08/02/2024	08/02/2024
							F1901	Additional Surgery	A24	No Apparent Device Problem	08/02/2024	08/02/2024
					CT314	Electrically powered	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	08/02/2024	08/02/2024
							F1901	Additional Surgery	A24	No Apparent Device Problem	08/02/2024	08/02/2024
					CT321	Long-term surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	08/02/2024	08/02/2024
							F1901	Additional Surgery	A24	No Apparent Device Problem	08/02/2024	08/02/2024
					CT449	Generators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	08/02/2024	08/02/2024
							F1901	Additional Surgery	A24	No Apparent Device Problem	08/02/2024	08/02/2024
					CT981	Single use	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	08/02/2024	08/02/2024
							F1901	Additional Surgery	A24	No Apparent Device Problem	08/02/2024	08/02/2024
					CT983	Surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	08/02/2024	08/02/2024
							F1901	Additional Surgery	A24	No Apparent Device Problem	08/02/2024	08/02/2024
14843	2024		47265	Dual-chamber implantable pacemaker, rate-responsive	CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	08/02/2024	08/02/2024
							F1901	Additional Surgery	A24	No Apparent Device Problem	08/02/2024	08/02/2024
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	08/02/2024	08/02/2024
							F1901	Additional Surgery	A24	No Apparent Device Problem	08/02/2024	08/02/2024
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	08/02/2024	08/02/2024
							F1901	Additional Surgery	A24	No Apparent Device Problem	08/02/2024	08/02/2024
14844	2024	All other reportable incidents	17882	Physiologic-monitoring defibrillation system	CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A0719	Unexpected Shutdown		
							F26	No Health Consequences or Impact	A0719	Unexpected Shutdown		
14845	2024	All other reportable incidents	47265	Dual-chamber implantable pacemaker, rate-responsive	CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	01/12/2023	01/12/2023
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	01/12/2023	01/12/2023
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	01/12/2023	01/12/2023
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	01/12/2023	01/12/2023
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	01/12/2023	01/12/2023
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	01/12/2023	01/12/2023
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	01/12/2023	01/12/2023
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	01/12/2023	01/12/2023
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	01/12/2023	01/12/2023
14846	2024		17882	Physiologic-monitoring defibrillation system	CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A0203	Defective Device		
									A0908	Incorrect, Inadequate or Imprecise Result or Readings		
									A090811	Unable to Obtain Readings		
							F26	No Health Consequences or Impact	A0203	Defective Device		
									A0908	Incorrect, Inadequate or Imprecise Result or Readings		
									A090811	Unable to Obtain Readings		
14847	2024	All other reportable incidents	47265	Dual-chamber implantable pacemaker, rate-responsive	CT1685	Cardiac pacemakers	E0112	Dizziness	A0712	Pacing Problem	19/09/2023	19/09/2023
									A13	Communication or Transmission Problem	19/09/2023	19/09/2023
							E0717	Dyspnea	A0712	Pacing Problem	19/09/2023	19/09/2023
									A13	Communication or Transmission Problem	19/09/2023	19/09/2023
							F1901	Additional Surgery	A0712	Pacing Problem	19/09/2023	19/09/2023
									A13	Communication or Transmission Problem	19/09/2023	19/09/2023
					CT257	Implantable cardiac pacemakers	E0112	Dizziness	A0712	Pacing Problem	19/09/2023	19/09/2023
									A13	Communication or Transmission Problem	19/09/2023	19/09/2023
							E0717	Dyspnea	A0712	Pacing Problem	19/09/2023	19/09/2023
									A13	Communication or Transmission Problem	19/09/2023	19/09/2023
							F1901	Additional Surgery	A0712	Pacing Problem	19/09/2023	19/09/2023
									A13	Communication or Transmission Problem	19/09/2023	19/09/2023
					CT2866	Implantable cardiac electrical stimulators	E0112	Dizziness	A0712	Pacing Problem	19/09/2023	19/09/2023
									A13	Communication or Transmission Problem	19/09/2023	19/09/2023
							E0717	Dyspnea	A0712	Pacing Problem	19/09/2023	19/09/2023
									A13	Communication or Transmission Problem	19/09/2023	19/09/2023
							F1901	Additional Surgery	A0712	Pacing Problem	19/09/2023	19/09/2023
									A13	Communication or Transmission Problem	19/09/2023	19/09/2023
14848	2024	All other reportable incidents	47267	Single-chamber implantable pacemaker, rate-responsive	CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	12/02/2024	12/02/2024
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	12/02/2024	12/02/2024
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	12/02/2024	12/02/2024
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	12/02/2024	12/02/2024
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	12/02/2024	12/02/2024

							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	12/02/2024	12/02/2024
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	12/02/2024	12/02/2024
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	12/02/2024	12/02/2024
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	12/02/2024	12/02/2024
14849	2024	All other reportable incidents	47270	Cardiac resynchronization therapy implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A070909	Oversensing		
									A0712	Pacing Problem		
									A072001	Electromagnetic Interference		
									A090801	Signal Artifact/Noise		
							F12	Serious Injury/ Illness/ Impairment	A070909	Oversensing		
									A0712	Pacing Problem		
									A072001	Electromagnetic Interference		
									A090801	Signal Artifact/Noise		
							F2307	Reprogramming of Device	A070909	Oversensing		
									A0712	Pacing Problem		
									A072001	Electromagnetic Interference		
									A090801	Signal Artifact/Noise		
14850	2024	All other reportable incidents	35223	Endocardial/interventricular septal pacing lead	CT1684	Cardiac pacemakers and associated devices	E0112	Dizziness	A0401	Break	07/02/2024	07/02/2024
									A070909	Oversensing	07/02/2024	07/02/2024
							F1901	Additional Surgery	A0401	Break	07/02/2024	07/02/2024
									A070909	Oversensing	07/02/2024	07/02/2024
							F1903	Device Explantation	A0401	Break	07/02/2024	07/02/2024
									A070909	Oversensing	07/02/2024	07/02/2024
							F1905	Device Revision or Replacement	A0401	Break	07/02/2024	07/02/2024
									A070909	Oversensing	07/02/2024	07/02/2024
14851	2024	All other reportable incidents	35223	Endocardial/interventricular septal pacing lead	CT1684	Cardiac pacemakers and associated devices	E0112	Dizziness	A0401	Break	07/02/2024	07/02/2024
									A070909	Oversensing	07/02/2024	07/02/2024
							F1901	Additional Surgery	A0401	Break	07/02/2024	07/02/2024
									A070909	Oversensing	07/02/2024	07/02/2024
14852	2024		47265	Dual-chamber implantable pacemaker, rate-responsive	CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	09/01/2024	09/01/2024
							F1901	Additional Surgery	A24	No Apparent Device Problem	09/01/2024	09/01/2024
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	09/01/2024	09/01/2024
							F1901	Additional Surgery	A24	No Apparent Device Problem	09/01/2024	09/01/2024
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	09/01/2024	09/01/2024
							F1901	Additional Surgery	A24	No Apparent Device Problem	09/01/2024	09/01/2024
14853	2024		47265	Dual-chamber implantable pacemaker, rate-responsive	CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	16/01/2024	16/01/2024
							F1901	Additional Surgery	A24	No Apparent Device Problem	16/01/2024	16/01/2024
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	16/01/2024	16/01/2024
							F1901	Additional Surgery	A24	No Apparent Device Problem	16/01/2024	16/01/2024
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	16/01/2024	16/01/2024
							F1901	Additional Surgery	A24	No Apparent Device Problem	16/01/2024	16/01/2024
14854	2024	All other reportable incidents	47267	Single-chamber implantable pacemaker, rate-responsive	CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A1205	Loose or Intermittent Connection		
							F26	No Health Consequences or Impact	A1205	Loose or Intermittent Connection		
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A1205	Loose or Intermittent Connection		
							F26	No Health Consequences or Impact	A1205	Loose or Intermittent Connection		
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A1205	Loose or Intermittent Connection		
							F26	No Health Consequences or Impact	A1205	Loose or Intermittent Connection		
14855	2024	All other reportable incidents	47265	Dual-chamber implantable pacemaker, rate-responsive	CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	23/02/2024	23/02/2024
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	23/02/2024	23/02/2024
							F12	Serious Injury/ Illness/ Impairment	A070504	Premature Discharge of Battery	23/02/2024	23/02/2024
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	23/02/2024	23/02/2024
							F23	Unexpected Medical Intervention	A070504	Premature Discharge of Battery	23/02/2024	23/02/2024
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	23/02/2024	23/02/2024
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	23/02/2024	23/02/2024
							F12	Serious Injury/ Illness/ Impairment	A070504	Premature Discharge of Battery	23/02/2024	23/02/2024
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	23/02/2024	23/02/2024
							F23	Unexpected Medical Intervention	A070504	Premature Discharge of Battery	23/02/2024	23/02/2024
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	23/02/2024	23/02/2024
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	23/02/2024	23/02/2024
							F12	Serious Injury/ Illness/ Impairment	A070504	Premature Discharge of Battery	23/02/2024	23/02/2024
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	23/02/2024	23/02/2024
							F23	Unexpected Medical Intervention	A070504	Premature Discharge of Battery	23/02/2024	23/02/2024
14856	2024	All other reportable incidents	35224	External pacemaker, epicardial pacing	CT1685	Cardiac pacemakers	E060109	Tachycardia	A0709	Device Sensing Problem		
							F1203	Life Threatening Illness or Injury	A0709	Device Sensing Problem		
14857	2024	All other reportable incidents	37265	Dual-chamber implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A070910	Undersensing		
									A071302	Inappropriate/Inadequate Shock/Stimulation		
							F12	Serious Injury/ Illness/ Impairment	A070910	Undersensing		
									A071302	Inappropriate/Inadequate Shock/Stimulation		
14858	2024		47265	Dual-chamber implantable pacemaker, rate-responsive	CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	06/02/2024	06/02/2024
							F1901	Additional Surgery	A24	No Apparent Device Problem	06/02/2024	06/02/2024
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	06/02/2024	06/02/2024
							F1901	Additional Surgery	A24	No Apparent Device Problem	06/02/2024	06/02/2024
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	06/02/2024	06/02/2024
							F1901	Additional Surgery	A24	No Apparent Device Problem	06/02/2024	06/02/2024
14859	2024	All other reportable incidents	47263	Cardiac resynchronization therapy implantable pacemaker	CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A1204	Failure to Disconnect	02/02/2024	02/02/2024
							F1901	Additional Surgery	A1204	Failure to Disconnect	02/02/2024	02/02/2024
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A1204	Failure to Disconnect	02/02/2024	02/02/2024
							F1901	Additional Surgery	A1204	Failure to Disconnect	02/02/2024	02/02/2024
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A1204	Failure to Disconnect	02/02/2024	02/02/2024
							F1901	Additional Surgery	A1204	Failure to Disconnect	02/02/2024	02/02/2024
14860	2024	All other reportable incidents	47263	Cardiac resynchronization therapy implantable pacemaker	CT1685	Cardiac pacemakers	E060101	Asystole	A070909	Oversensing		
									A0712	Pacing Problem		
									A090801	Signal Artifact/Noise		
							F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A070909	Oversensing		
									A0712	Pacing Problem		
									A090801	Signal Artifact/Noise		
					CT257	Implantable cardiac pacemakers	E060101	Asystole	A070909	Oversensing		
									A0712	Pacing Problem		
									A090801	Signal Artifact/Noise		

							F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A070909	Oversensing		
									A0712	Pacing Problem		
									A090801	Signal Artifact/Noise		
					CT2866	Implantable cardiac electrical stimulators	E060101	Asystole	A070909	Oversensing		
									A0712	Pacing Problem		
									A090801	Signal Artifact/Noise		
							F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A070909	Oversensing		
									A0712	Pacing Problem		
									A090801	Signal Artifact/Noise		
14861	2024	All other reportable incidents	47265	Dual-chamber implantable pacemaker, rate-responsive	CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	01/02/2024	01/02/2024
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	01/02/2024	01/02/2024
							F1903	Device Explantation	A070504	Premature Discharge of Battery	01/02/2024	01/02/2024
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	01/02/2024	01/02/2024
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	01/02/2024	01/02/2024
							F1903	Device Explantation	A070504	Premature Discharge of Battery	01/02/2024	01/02/2024
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	01/02/2024	01/02/2024
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	01/02/2024	01/02/2024
							F1903	Device Explantation	A070504	Premature Discharge of Battery	01/02/2024	01/02/2024
14862	2024		47265	Dual-chamber implantable pacemaker, rate-responsive	CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A25	No Apparent Adverse Event		
							F26	No Health Consequences or Impact	A25	No Apparent Adverse Event		
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A25	No Apparent Adverse Event		
							F26	No Health Consequences or Impact	A25	No Apparent Adverse Event		
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A25	No Apparent Adverse Event		
							F26	No Health Consequences or Impact	A25	No Apparent Adverse Event		
14863	2024		47265	Dual-chamber implantable pacemaker, rate-responsive	CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	09/02/2024	09/02/2024
							F1901	Additional Surgery	A24	No Apparent Device Problem	09/02/2024	09/02/2024
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	09/02/2024	09/02/2024
							F1901	Additional Surgery	A24	No Apparent Device Problem	09/02/2024	09/02/2024
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	09/02/2024	09/02/2024
							F1901	Additional Surgery	A24	No Apparent Device Problem	09/02/2024	09/02/2024
14864	2024		47265	Dual-chamber implantable pacemaker, rate-responsive	CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem		
							F1901	Additional Surgery	A24	No Apparent Device Problem		
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem		
							F1901	Additional Surgery	A24	No Apparent Device Problem		
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem		
							F1901	Additional Surgery	A24	No Apparent Device Problem		
14865	2024	All other reportable incidents	35852	Single-chamber implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	06/09/2023	06/09/2023
									A0908	Incorrect, Inadequate or Imprecise Result or Readings	06/09/2023	06/09/2023
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	06/09/2023	06/09/2023
									A0908	Incorrect, Inadequate or Imprecise Result or Readings	06/09/2023	06/09/2023
							F1903	Device Explantation	A070504	Premature Discharge of Battery	06/09/2023	06/09/2023
									A0908	Incorrect, Inadequate or Imprecise Result or Readings	06/09/2023	06/09/2023
14866	2024	All other reportable incidents	35852	Single-chamber implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	28/08/2023	28/08/2023
									A0908	Incorrect, Inadequate or Imprecise Result or Readings	28/08/2023	28/08/2023
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	28/08/2023	28/08/2023
									A0908	Incorrect, Inadequate or Imprecise Result or Readings	28/08/2023	28/08/2023
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	28/08/2023	28/08/2023
									A0908	Incorrect, Inadequate or Imprecise Result or Readings	28/08/2023	28/08/2023
14867	2024	All other reportable incidents	17882	Physiologic-monitoring defibrillation system	CT1684	Cardiac pacemakers and associated devices			A071301	Failure to Deliver Shock/Stimulation		
14868	2024	All other reportable incidents	17882	Physiologic-monitoring defibrillation system	CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A0708	Power Problem		
							F27	Problem Identified During Non-Clinical Procedure	A0708	Power Problem		
14869	2024	All other reportable incidents	17882	Physiologic-monitoring defibrillation system	CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A071204	No Pacing		
							F27	Problem Identified During Non-Clinical Procedure	A071204	No Pacing		
14870	2024	All other reportable incidents	17882	Physiologic-monitoring defibrillation system	CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A071204	No Pacing		
							F27	Problem Identified During Non-Clinical Procedure	A071204	No Pacing		
14871	2024	All other reportable incidents	47263	Cardiac resynchronization therapy implantable pacemaker	CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A070909	Oversensing	05/10/2023	05/10/2023
									A0712	Pacing Problem	05/10/2023	05/10/2023
									A0908	Incorrect, Inadequate or Imprecise Result or Readings	05/10/2023	05/10/2023
									A090801	Signal Artifact/Noise	05/10/2023	05/10/2023
							F08	Hospitalization or Prolonged Hospitalization	A070909	Oversensing	05/10/2023	05/10/2023
									A0712	Pacing Problem	05/10/2023	05/10/2023
									A0908	Incorrect, Inadequate or Imprecise Result or Readings	05/10/2023	05/10/2023
									A090801	Signal Artifact/Noise	05/10/2023	05/10/2023
							F1001	Absence of Treatment	A070909	Oversensing	05/10/2023	05/10/2023
									A0712	Pacing Problem	05/10/2023	05/10/2023
									A0908	Incorrect, Inadequate or Imprecise Result or Readings	05/10/2023	05/10/2023
									A090801	Signal Artifact/Noise	05/10/2023	05/10/2023
							F1905	Device Revision or Replacement	A070909	Oversensing	05/10/2023	05/10/2023
									A0712	Pacing Problem	05/10/2023	05/10/2023
									A0908	Incorrect, Inadequate or Imprecise Result or Readings	05/10/2023	05/10/2023
									A090801	Signal Artifact/Noise	05/10/2023	05/10/2023
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A070909	Oversensing	05/10/2023	05/10/2023
									A0712	Pacing Problem	05/10/2023	05/10/2023
									A0908	Incorrect, Inadequate or Imprecise Result or Readings	05/10/2023	05/10/2023
									A090801	Signal Artifact/Noise	05/10/2023	05/10/2023
							F08	Hospitalization or Prolonged Hospitalization	A070909	Oversensing	05/10/2023	05/10/2023
									A0712	Pacing Problem	05/10/2023	05/10/2023
									A0908	Incorrect, Inadequate or Imprecise Result or Readings	05/10/2023	05/10/2023
									A090801	Signal Artifact/Noise	05/10/2023	05/10/2023

						F1001	Absence of Treatment	A070909	Oversensing	05/10/2023	05/10/2023
								A0712	Pacing Problem	05/10/2023	05/10/2023
								A0908	Incorrect, Inadequate or Imprecise Result or Readings	05/10/2023	05/10/2023
								A090801	Signal Artifact/Noise	05/10/2023	05/10/2023
						F1905	Device Revision or Replacement	A070909	Oversensing	05/10/2023	05/10/2023
								A0712	Pacing Problem	05/10/2023	05/10/2023
								A0908	Incorrect, Inadequate or Imprecise Result or Readings	05/10/2023	05/10/2023
								A090801	Signal Artifact/Noise	05/10/2023	05/10/2023
				CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A070909	Oversensing	05/10/2023	05/10/2023
								A0712	Pacing Problem	05/10/2023	05/10/2023
								A0908	Incorrect, Inadequate or Imprecise Result or Readings	05/10/2023	05/10/2023
								A090801	Signal Artifact/Noise	05/10/2023	05/10/2023
						F08	Hospitalization or Prolonged Hospitalization	A070909	Oversensing	05/10/2023	05/10/2023
								A0712	Pacing Problem	05/10/2023	05/10/2023
								A0908	Incorrect, Inadequate or Imprecise Result or Readings	05/10/2023	05/10/2023
								A090801	Signal Artifact/Noise	05/10/2023	05/10/2023
						F1001	Absence of Treatment	A070909	Oversensing	05/10/2023	05/10/2023
								A0712	Pacing Problem	05/10/2023	05/10/2023
								A0908	Incorrect, Inadequate or Imprecise Result or Readings	05/10/2023	05/10/2023
								A090801	Signal Artifact/Noise	05/10/2023	05/10/2023
						F1905	Device Revision or Replacement	A070909	Oversensing	05/10/2023	05/10/2023
								A0712	Pacing Problem	05/10/2023	05/10/2023
								A0908	Incorrect, Inadequate or Imprecise Result or Readings	05/10/2023	05/10/2023
								A090801	Signal Artifact/Noise	05/10/2023	05/10/2023
14872	2024	All other reportable incidents	47265	Dual-chamber implantable pacemaker, rate-responsive	CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
14873	2024	All other reportable incidents	37265	Dual-chamber implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A072202	Low Impedance		
									A1601	Device Alarm System		
							F26	No Health Consequences or Impact	A072202	Low Impedance		
									A1601	Device Alarm System		
14874	2024	All other reportable incidents	47205	Cardiac pulse generator programmer	CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A1102	Application Program Problem		
									A150105	Key or Button Unresponsive/not Working		
							F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A1102	Application Program Problem		
									A150105	Key or Button Unresponsive/not Working		
14875	2024	All other reportable incidents	35852	Single-chamber implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E060109	Tachycardia	A0709	Device Sensing Problem		
									A070909	Oversensing		
									A072001	Electromagnetic Interference		
									A090801	Signal Artifact/Noise		
							F26	No Health Consequences or Impact	A0709	Device Sensing Problem		
									A070909	Oversensing		
									A072001	Electromagnetic Interference		
									A090801	Signal Artifact/Noise		
14876	2024	All other reportable incidents	47265	Dual-chamber implantable pacemaker, rate-responsive	CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A070909	Oversensing	06/06/2024	06/06/2024
									A071204	No Pacing	06/06/2024	06/06/2024
							F1901	Additional Surgery	A070909	Oversensing	06/06/2024	06/06/2024
									A071204	No Pacing	06/06/2024	06/06/2024
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A070909	Oversensing	06/06/2024	06/06/2024
									A071204	No Pacing	06/06/2024	06/06/2024
							F1901	Additional Surgery	A070909	Oversensing	06/06/2024	06/06/2024
									A071204	No Pacing	06/06/2024	06/06/2024
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A070909	Oversensing	06/06/2024	06/06/2024
									A071204	No Pacing	06/06/2024	06/06/2024
							F1901	Additional Surgery	A070909	Oversensing	06/06/2024	06/06/2024
									A071204	No Pacing	06/06/2024	06/06/2024
14877	2024	All other reportable incidents	47263	Cardiac resynchronization therapy implantable pacemaker	CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A0908	Incorrect, Inadequate or Imprecise Result or Readings	22/12/2023	22/12/2023
							F08	Hospitalization or Prolonged Hospitalization	A0908	Incorrect, Inadequate or Imprecise Result or Readings	22/12/2023	22/12/2023
							F1905	Device Revision or Replacement	A0908	Incorrect, Inadequate or Imprecise Result or Readings	22/12/2023	22/12/2023
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A0908	Incorrect, Inadequate or Imprecise Result or Readings	22/12/2023	22/12/2023
							F08	Hospitalization or Prolonged Hospitalization	A0908	Incorrect, Inadequate or Imprecise Result or Readings	22/12/2023	22/12/2023
							F1905	Device Revision or Replacement	A0908	Incorrect, Inadequate or Imprecise Result or Readings	22/12/2023	22/12/2023
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A0908	Incorrect, Inadequate or Imprecise Result or Readings	22/12/2023	22/12/2023
							F08	Hospitalization or Prolonged Hospitalization	A0908	Incorrect, Inadequate or Imprecise Result or Readings	22/12/2023	22/12/2023
							F1905	Device Revision or Replacement	A0908	Incorrect, Inadequate or Imprecise Result or Readings	22/12/2023	22/12/2023
14878	2024	All other reportable incidents	35852	Single-chamber implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E2104	Electric Shock	A070909	Oversensing	02/11/2023	02/11/2023
									A071302	Inappropriate/Inadequate Shock/Stimulation	02/11/2023	02/11/2023
									A090801	Signal Artifact/Noise	02/11/2023	02/11/2023

								A1301	Failure to Read Input Signal	02/11/2023	02/11/2023
						F08	Hospitalization or Prolonged Hospitalization	A070909	Oversensing	02/11/2023	02/11/2023
								A071302	Inappropriate/Inadequate Shock/Stimulation	02/11/2023	02/11/2023
								A090801	Signal Artifact/Noise	02/11/2023	02/11/2023
								A1301	Failure to Read Input Signal	02/11/2023	02/11/2023
						F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A070909	Oversensing	02/11/2023	02/11/2023
								A071302	Inappropriate/Inadequate Shock/Stimulation	02/11/2023	02/11/2023
								A090801	Signal Artifact/Noise	02/11/2023	02/11/2023
								A1301	Failure to Read Input Signal	02/11/2023	02/11/2023
						F1905	Device Revision or Replacement	A070909	Oversensing	02/11/2023	02/11/2023
								A071302	Inappropriate/Inadequate Shock/Stimulation	02/11/2023	02/11/2023
								A090801	Signal Artifact/Noise	02/11/2023	02/11/2023
								A1301	Failure to Read Input Signal	02/11/2023	02/11/2023
14879	2024	All other reportable incidents	60789	Intracardiac pacemaker	CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A070101	Failure to Capture		
							F23	Unexpected Medical Intervention	A070101	Failure to Capture		
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A070101	Failure to Capture		
							F23	Unexpected Medical Intervention	A070101	Failure to Capture		
					CT2818	Catheterization/Cannulation/Fluid path devices	E2403	No Clinical Signs, Symptoms or Conditions	A070101	Failure to Capture		
							F23	Unexpected Medical Intervention	A070101	Failure to Capture		
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A070101	Failure to Capture		
							F23	Unexpected Medical Intervention	A070101	Failure to Capture		
14880	2024	All other reportable incidents	47265	Dual-chamber implantable pacemaker, rate-responsive	CT1685	Cardiac pacemakers	E0112	Dizziness	A070909	Oversensing	01/01/2024	01/01/2024
									A0712	Pacing Problem	01/01/2024	01/01/2024
									A0908	Incorrect, Inadequate or Imprecise Result or Readings	01/01/2024	01/01/2024
							E060104	Bradycardia	A070909	Oversensing	01/01/2024	01/01/2024
									A0712	Pacing Problem	01/01/2024	01/01/2024
									A0908	Incorrect, Inadequate or Imprecise Result or Readings	01/01/2024	01/01/2024
							F08	Hospitalization or Prolonged Hospitalization	A070909	Oversensing	01/01/2024	01/01/2024
									A0712	Pacing Problem	01/01/2024	01/01/2024
									A0908	Incorrect, Inadequate or Imprecise Result or Readings	01/01/2024	01/01/2024
							F1001	Absence of Treatment	A070909	Oversensing	01/01/2024	01/01/2024
									A0712	Pacing Problem	01/01/2024	01/01/2024
									A0908	Incorrect, Inadequate or Imprecise Result or Readings	01/01/2024	01/01/2024
							F1905	Device Revision or Replacement	A070909	Oversensing	01/01/2024	01/01/2024
									A0712	Pacing Problem	01/01/2024	01/01/2024
									A0908	Incorrect, Inadequate or Imprecise Result or Readings	01/01/2024	01/01/2024
					CT257	Implantable cardiac pacemakers	E0112	Dizziness	A070909	Oversensing	01/01/2024	01/01/2024
									A0712	Pacing Problem	01/01/2024	01/01/2024
									A0908	Incorrect, Inadequate or Imprecise Result or Readings	01/01/2024	01/01/2024
							E060104	Bradycardia	A070909	Oversensing	01/01/2024	01/01/2024
									A0712	Pacing Problem	01/01/2024	01/01/2024
									A0908	Incorrect, Inadequate or Imprecise Result or Readings	01/01/2024	01/01/2024
							F08	Hospitalization or Prolonged Hospitalization	A070909	Oversensing	01/01/2024	01/01/2024
									A0712	Pacing Problem	01/01/2024	01/01/2024
									A0908	Incorrect, Inadequate or Imprecise Result or Readings	01/01/2024	01/01/2024
							F1001	Absence of Treatment	A070909	Oversensing	01/01/2024	01/01/2024
									A0712	Pacing Problem	01/01/2024	01/01/2024
									A0908	Incorrect, Inadequate or Imprecise Result or Readings	01/01/2024	01/01/2024
							F1905	Device Revision or Replacement	A070909	Oversensing	01/01/2024	01/01/2024
									A0712	Pacing Problem	01/01/2024	01/01/2024
									A0908	Incorrect, Inadequate or Imprecise Result or Readings	01/01/2024	01/01/2024
					CT2866	Implantable cardiac electrical stimulators	E0112	Dizziness	A070909	Oversensing	01/01/2024	01/01/2024
									A0712	Pacing Problem	01/01/2024	01/01/2024
									A0908	Incorrect, Inadequate or Imprecise Result or Readings	01/01/2024	01/01/2024
							E060104	Bradycardia	A070909	Oversensing	01/01/2024	01/01/2024
									A0712	Pacing Problem	01/01/2024	01/01/2024
									A0908	Incorrect, Inadequate or Imprecise Result or Readings	01/01/2024	01/01/2024
							F08	Hospitalization or Prolonged Hospitalization	A070909	Oversensing	01/01/2024	01/01/2024
									A0712	Pacing Problem	01/01/2024	01/01/2024
									A0908	Incorrect, Inadequate or Imprecise Result or Readings	01/01/2024	01/01/2024
							F1001	Absence of Treatment	A070909	Oversensing	01/01/2024	01/01/2024
									A0712	Pacing Problem	01/01/2024	01/01/2024
									A0908	Incorrect, Inadequate or Imprecise Result or Readings	01/01/2024	01/01/2024
							F1905	Device Revision or Replacement	A070909	Oversensing	01/01/2024	01/01/2024
									A0712	Pacing Problem	01/01/2024	01/01/2024
									A0908	Incorrect, Inadequate or Imprecise Result or Readings	01/01/2024	01/01/2024
14881	2024	All other reportable incidents	47265	Dual-chamber implantable pacemaker, rate-responsive	CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A071102	Failure to Interrogate		
							F26	No Health Consequences or Impact	A071102	Failure to Interrogate		
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A071102	Failure to Interrogate		
							F26	No Health Consequences or Impact	A071102	Failure to Interrogate		
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A071102	Failure to Interrogate		
							F26	No Health Consequences or Impact	A071102	Failure to Interrogate		
14882	2024	All other reportable incidents	47265	Dual-chamber implantable pacemaker, rate-responsive	CT1685	Cardiac pacemakers	E2103	Device Overstimulation of Tissue	A0908	Incorrect, Inadequate or Imprecise Result or Readings	01/08/2024	01/08/2024
							F08	Hospitalization or Prolonged Hospitalization	A0908	Incorrect, Inadequate or Imprecise Result or Readings	01/08/2024	01/08/2024

							F1905	Device Revision or Replacement	A0908	Incorrect, Inadequate or Imprecise Result or Readings	01/08/2024	01/08/2024
					CT257	Implantable cardiac pacemakers	E2103	Device Overstimulation of Tissue	A0908	Incorrect, Inadequate or Imprecise Result or Readings	01/08/2024	01/08/2024
							F08	Hospitalization or Prolonged Hospitalization	A0908	Incorrect, Inadequate or Imprecise Result or Readings	01/08/2024	01/08/2024
							F1905	Device Revision or Replacement	A0908	Incorrect, Inadequate or Imprecise Result or Readings	01/08/2024	01/08/2024
					CT2866	Implantable cardiac electrical stimulators	E2103	Device Overstimulation of Tissue	A0908	Incorrect, Inadequate or Imprecise Result or Readings	01/08/2024	01/08/2024
							F08	Hospitalization or Prolonged Hospitalization	A0908	Incorrect, Inadequate or Imprecise Result or Readings	01/08/2024	01/08/2024
							F1905	Device Revision or Replacement	A0908	Incorrect, Inadequate or Imprecise Result or Readings	01/08/2024	01/08/2024
14883	2024	All other reportable incidents	37265	Dual-chamber implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem		
							F26	No Health Consequences or Impact	A0705	Battery Problem		
14884	2024	All other reportable incidents	35852	Single-chamber implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A1601	Device Alarm System		
							F26	No Health Consequences or Impact	A1601	Device Alarm System		
14885	2024	All other reportable incidents	35852	Single-chamber implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A090813	Imprecise or Erratic Results		
							F27	Problem Identified During Non-Clinical Procedure	A090813	Imprecise or Erratic Results		
14886	2024	All other reportable incidents	17882	Physiologic-monitoring defibrillation system	CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A071301	Failure to Deliver Shock/Stimulation		
							F27	Problem Identified During Non-Clinical Procedure	A071301	Failure to Deliver Shock/Stimulation		
14887	2024	All other reportable incidents	17882	Physiologic-monitoring defibrillation system	CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A150105	Key or Button Unresponsive/not Working		
							F27	Problem Identified During Non-Clinical Procedure	A150105	Key or Button Unresponsive/not Working		
14888	2024	All other reportable incidents	17882	Physiologic-monitoring defibrillation system	CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070912	Incorrect Interpretation of Signal		
							F26	No Health Consequences or Impact	A070912	Incorrect Interpretation of Signal		
14889	2024		47265	Dual-chamber implantable pacemaker, rate-responsive	CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A25	No Apparent Adverse Event		
							F1905	Device Revision or Replacement	A25	No Apparent Adverse Event		
							F22	Unexpected Diagnostic Intervention	A25	No Apparent Adverse Event		
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A25	No Apparent Adverse Event		
							F1905	Device Revision or Replacement	A25	No Apparent Adverse Event		
							F22	Unexpected Diagnostic Intervention	A25	No Apparent Adverse Event		
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A25	No Apparent Adverse Event		
							F1905	Device Revision or Replacement	A25	No Apparent Adverse Event		
							F22	Unexpected Diagnostic Intervention	A25	No Apparent Adverse Event		
14890	2024		35224	External pacemaker, epicardial pacing	CT1685	Cardiac pacemakers	E0602	Cardiac Arrest	A01	Patient Device Interaction Problem		
									A050502	Misfire		
									A07	Electrical /Electronic Property Problem		
									A0709	Device Sensing Problem		
							F1203	Life Threatening Illness or Injury	A01	Patient Device Interaction Problem		
									A050502	Misfire		
									A07	Electrical /Electronic Property Problem		
									A0709	Device Sensing Problem		
14891	2024	All other reportable incidents	47270	Cardiac resynchronization therapy implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A0709	Device Sensing Problem		
									A070909	Oversensing		
									A070910	Undersensing		
									A0712	Pacing Problem		
									A0722	Impedance Problem		
									A1301	Failure to Read Input Signal		
							F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A0709	Device Sensing Problem		
									A070909	Oversensing		
									A070910	Undersensing		
									A0712	Pacing Problem		
									A0722	Impedance Problem		
									A1301	Failure to Read Input Signal		
14892	2024	All other reportable incidents	47265	Dual-chamber implantable pacemaker, rate-responsive	CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	25/06/2024	25/06/2024
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	25/06/2024	25/06/2024
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	25/06/2024	25/06/2024
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	25/06/2024	25/06/2024
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	25/06/2024	25/06/2024
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	25/06/2024	25/06/2024
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	25/06/2024	25/06/2024
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	25/06/2024	25/06/2024
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	25/06/2024	25/06/2024
14893	2024	All other reportable incidents	35852	Single-chamber implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem	22/11/2023	22/11/2023
									A070504	Premature Discharge of Battery	22/11/2023	22/11/2023
							F08	Hospitalization or Prolonged Hospitalization	A0705	Battery Problem	22/11/2023	22/11/2023
									A070504	Premature Discharge of Battery	22/11/2023	22/11/2023
							F1903	Device Explantation	A0705	Battery Problem	22/11/2023	22/11/2023
									A070504	Premature Discharge of Battery	22/11/2023	22/11/2023
							F1905	Device Revision or Replacement	A0705	Battery Problem	22/11/2023	22/11/2023
									A070504	Premature Discharge of Battery	22/11/2023	22/11/2023
14894	2024	All other reportable incidents	35852	Single-chamber implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	20/10/2023	20/10/2023
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	20/10/2023	20/10/2023
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	20/10/2023	20/10/2023
14895	2024	All other reportable incidents	35974	Cardiac electrophysiology stimulation system	CT1684	Cardiac pacemakers and associated devices	E0601	Arrhythmia	A110201	Application Program Freezes or Fails to Launch		
							F05	Delay to Treatment/ Therapy	A110201	Application Program Freezes or Fails to Launch		
14896	2024	All other reportable incidents	35974	Cardiac electrophysiology stimulation system	CT1684	Cardiac pacemakers and associated devices	E0601	Arrhythmia	A090203	Erratic or Intermittent Display		
							F05	Delay to Treatment/ Therapy	A090203	Erratic or Intermittent Display		
14897	2024	All other reportable incidents	47265	Dual-chamber implantable pacemaker, rate-responsive	CT1685	Cardiac pacemakers	E2401	Insufficient Information	A071102	Failure to Interrogate		
							F02	Death	A071102	Failure to Interrogate		
					CT257	Implantable cardiac pacemakers	E2401	Insufficient Information	A071102	Failure to Interrogate		
							F02	Death	A071102	Failure to Interrogate		
					CT2866	Implantable cardiac electrical stimulators	E2401	Insufficient Information	A071102	Failure to Interrogate		
							F02	Death	A071102	Failure to Interrogate		
14898	2024	All other reportable incidents	35852	Single-chamber implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A072201	High Impedance		
							F26	No Health Consequences or Impact	A072201	High Impedance		
14899	2024	All other reportable incidents	37265	Dual-chamber implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A070909	Oversensing		
									A0712	Pacing Problem		

							F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A070909	Oversensing		
									A0712	Pacing Problem		
14900	2024		47265	Dual-chamber implantable pacemaker, rate-responsive	CT1685	Cardiac pacemakers	E0119	Loss of Consciousness	A160602	Premature End-of-Life Indicator		
							F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A160602	Premature End-of-Life Indicator		
					CT257	Implantable cardiac pacemakers	E0119	Loss of Consciousness	A160602	Premature End-of-Life Indicator		
							F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A160602	Premature End-of-Life Indicator		
					CT2866	Implantable cardiac electrical stimulators	E0119	Loss of Consciousness	A160602	Premature End-of-Life Indicator		
							F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A160602	Premature End-of-Life Indicator		
14901	2024	All other reportable incidents	17882	Physiologic-monitoring defibrillation system	CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A07	Electrical /Electronic Property Problem		
							F27	Problem Identified During Non-Clinical Procedure	A07	Electrical /Electronic Property Problem		
14902	2024		47265	Dual-chamber implantable pacemaker, rate-responsive	CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem		
							F1901	Additional Surgery	A24	No Apparent Device Problem		
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem		
							F1901	Additional Surgery	A24	No Apparent Device Problem		
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem		
							F1901	Additional Surgery	A24	No Apparent Device Problem		
14903	2024	All other reportable incidents	47265	Dual-chamber implantable pacemaker, rate-responsive	CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	20/02/2024	20/02/2024
							F1901	Additional Surgery	A070504	Premature Discharge of Battery	20/02/2024	20/02/2024
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	20/02/2024	20/02/2024
							F1901	Additional Surgery	A070504	Premature Discharge of Battery	20/02/2024	20/02/2024
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	20/02/2024	20/02/2024
							F1901	Additional Surgery	A070504	Premature Discharge of Battery	20/02/2024	20/02/2024
14904	2024		47265	Dual-chamber implantable pacemaker, rate-responsive	CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F1903	Device Explantation	A070504	Premature Discharge of Battery		
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F1903	Device Explantation	A070504	Premature Discharge of Battery		
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F1903	Device Explantation	A070504	Premature Discharge of Battery		
14905	2024	All other reportable incidents	47263	Cardiac resynchronization therapy implantable pacemaker	CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A090201	Device Displays Incorrect Message	13/02/2024	13/02/2024
							F08	Hospitalization or Prolonged Hospitalization	A090201	Device Displays Incorrect Message	13/02/2024	13/02/2024
							F1905	Device Revision or Replacement	A090201	Device Displays Incorrect Message	13/02/2024	13/02/2024
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A090201	Device Displays Incorrect Message	13/02/2024	13/02/2024
							F08	Hospitalization or Prolonged Hospitalization	A090201	Device Displays Incorrect Message	13/02/2024	13/02/2024
							F1905	Device Revision or Replacement	A090201	Device Displays Incorrect Message	13/02/2024	13/02/2024
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A090201	Device Displays Incorrect Message	13/02/2024	13/02/2024
							F08	Hospitalization or Prolonged Hospitalization	A090201	Device Displays Incorrect Message	13/02/2024	13/02/2024
							F1905	Device Revision or Replacement	A090201	Device Displays Incorrect Message	13/02/2024	13/02/2024
14906	2024	All other reportable incidents	35852	Single-chamber implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	12/12/2023	12/12/2023
									A0908	Incorrect, Inadequate or Imprecise Result or Readings	12/12/2023	12/12/2023
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	12/12/2023	12/12/2023
									A0908	Incorrect, Inadequate or Imprecise Result or Readings	12/12/2023	12/12/2023
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	12/12/2023	12/12/2023
									A0908	Incorrect, Inadequate or Imprecise Result or Readings	12/12/2023	12/12/2023
14907	2024	All other reportable incidents	47263	Cardiac resynchronization therapy implantable pacemaker	CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A0908	Incorrect, Inadequate or Imprecise Result or Readings	15/12/2023	15/12/2023
							F08	Hospitalization or Prolonged Hospitalization	A0908	Incorrect, Inadequate or Imprecise Result or Readings	15/12/2023	15/12/2023
							F1905	Device Revision or Replacement	A0908	Incorrect, Inadequate or Imprecise Result or Readings	15/12/2023	15/12/2023
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A0908	Incorrect, Inadequate or Imprecise Result or Readings	15/12/2023	15/12/2023
							F08	Hospitalization or Prolonged Hospitalization	A0908	Incorrect, Inadequate or Imprecise Result or Readings	15/12/2023	15/12/2023
							F1905	Device Revision or Replacement	A0908	Incorrect, Inadequate or Imprecise Result or Readings	15/12/2023	15/12/2023
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A0908	Incorrect, Inadequate or Imprecise Result or Readings	15/12/2023	15/12/2023
							F08	Hospitalization or Prolonged Hospitalization	A0908	Incorrect, Inadequate or Imprecise Result or Readings	15/12/2023	15/12/2023
							F1905	Device Revision or Replacement	A0908	Incorrect, Inadequate or Imprecise Result or Readings	15/12/2023	15/12/2023
14908	2024	All other reportable incidents	35852	Single-chamber implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
14909	2024		47265	Dual-chamber implantable pacemaker, rate-responsive	CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	08/02/2024	08/02/2024
							F1901	Additional Surgery	A24	No Apparent Device Problem	08/02/2024	08/02/2024
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	08/02/2024	08/02/2024
							F1901	Additional Surgery	A24	No Apparent Device Problem	08/02/2024	08/02/2024
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	08/02/2024	08/02/2024
							F1901	Additional Surgery	A24	No Apparent Device Problem	08/02/2024	08/02/2024
14910	2024		47265	Dual-chamber implantable pacemaker, rate-responsive	CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	04/01/2024	04/01/2024
							F1901	Additional Surgery	A24	No Apparent Device Problem	04/01/2024	04/01/2024
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	04/01/2024	04/01/2024
							F1901	Additional Surgery	A24	No Apparent Device Problem	04/01/2024	04/01/2024
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	04/01/2024	04/01/2024
							F1901	Additional Surgery	A24	No Apparent Device Problem	04/01/2024	04/01/2024
14911	2024		47265	Dual-chamber implantable pacemaker, rate-responsive	CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	29/01/2024	29/01/2024
							F1901	Additional Surgery	A24	No Apparent Device Problem	29/01/2024	29/01/2024
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	29/01/2024	29/01/2024
							F1901	Additional Surgery	A24	No Apparent Device Problem	29/01/2024	29/01/2024
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	29/01/2024	29/01/2024
							F1901	Additional Surgery	A24	No Apparent Device Problem	29/01/2024	29/01/2024
14912	2024		47265	Dual-chamber implantable pacemaker, rate-responsive	CT141	Electrodes/Leads and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	31/01/2024	31/01/2024
							F1901	Additional Surgery	A24	No Apparent Device Problem	31/01/2024	31/01/2024
					CT145	Cardiovascular devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	31/01/2024	31/01/2024
							F1901	Additional Surgery	A24	No Apparent Device Problem	31/01/2024	31/01/2024
					CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	31/01/2024	31/01/2024

							F1901	Additional Surgery	A24	No Apparent Device Problem	31/01/2024	31/01/2024
					CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	31/01/2024	31/01/2024
							F1901	Additional Surgery	A24	No Apparent Device Problem	31/01/2024	31/01/2024
					CT233	Surgical	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	31/01/2024	31/01/2024
							F1901	Additional Surgery	A24	No Apparent Device Problem	31/01/2024	31/01/2024
					CT2406	Implantable	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	31/01/2024	31/01/2024
							F1901	Additional Surgery	A24	No Apparent Device Problem	31/01/2024	31/01/2024
					CT2407	Active implantable	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	31/01/2024	31/01/2024
							F1901	Additional Surgery	A24	No Apparent Device Problem	31/01/2024	31/01/2024
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	31/01/2024	31/01/2024
							F1901	Additional Surgery	A24	No Apparent Device Problem	31/01/2024	31/01/2024
					CT2727	Active implant devices/ancillaries	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	31/01/2024	31/01/2024
							F1901	Additional Surgery	A24	No Apparent Device Problem	31/01/2024	31/01/2024
					CT2750	Implantable cardiac electrical stimulators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	31/01/2024	31/01/2024
							F1901	Additional Surgery	A24	No Apparent Device Problem	31/01/2024	31/01/2024
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	31/01/2024	31/01/2024
							F1901	Additional Surgery	A24	No Apparent Device Problem	31/01/2024	31/01/2024
					CT2985	Active	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	31/01/2024	31/01/2024
							F1901	Additional Surgery	A24	No Apparent Device Problem	31/01/2024	31/01/2024
					CT3001	Implantable metal	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	31/01/2024	31/01/2024
							F1901	Additional Surgery	A24	No Apparent Device Problem	31/01/2024	31/01/2024
					CT3002	Implantable synthetic polymer	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	31/01/2024	31/01/2024
							F1901	Additional Surgery	A24	No Apparent Device Problem	31/01/2024	31/01/2024
					CT3006	Implantable inorganic material	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	31/01/2024	31/01/2024
							F1901	Additional Surgery	A24	No Apparent Device Problem	31/01/2024	31/01/2024
					CT314	Electrically powered	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	31/01/2024	31/01/2024
							F1901	Additional Surgery	A24	No Apparent Device Problem	31/01/2024	31/01/2024
					CT321	Long-term surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	31/01/2024	31/01/2024
							F1901	Additional Surgery	A24	No Apparent Device Problem	31/01/2024	31/01/2024
					CT449	Generators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	31/01/2024	31/01/2024
							F1901	Additional Surgery	A24	No Apparent Device Problem	31/01/2024	31/01/2024
					CT981	Single use	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	31/01/2024	31/01/2024
							F1901	Additional Surgery	A24	No Apparent Device Problem	31/01/2024	31/01/2024
					CT983	Surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	31/01/2024	31/01/2024
							F1901	Additional Surgery	A24	No Apparent Device Problem	31/01/2024	31/01/2024
14913	2024	All other reportable incidents	17882	Physiologic-monitoring defibrillation system	CT1684	Cardiac pacemakers and associated devices	E0602	Cardiac Arrest	A0719	Unexpected Shutdown		
							F26	No Health Consequences or Impact	A0719	Unexpected Shutdown		
14914	2024	All other reportable incidents	37265	Dual-chamber implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E060109	Tachycardia	A0709	Device Sensing Problem		
							F1203	Life Threatening Illness or Injury	A0709	Device Sensing Problem		
							F2307	Reprogramming of Device	A0709	Device Sensing Problem		
14915	2024		47265	Dual-chamber implantable pacemaker, rate-responsive	CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	15/01/2024	15/01/2024
							F1901	Additional Surgery	A24	No Apparent Device Problem	15/01/2024	15/01/2024
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	15/01/2024	15/01/2024
							F1901	Additional Surgery	A24	No Apparent Device Problem	15/01/2024	15/01/2024
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	15/01/2024	15/01/2024
							F1901	Additional Surgery	A24	No Apparent Device Problem	15/01/2024	15/01/2024
14916	2024	All other reportable incidents	35852	Single-chamber implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
14917	2024		47267	Single-chamber implantable pacemaker, rate-responsive	CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
14918	2024	All other reportable incidents	47267	Single-chamber implantable pacemaker, rate-responsive	CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	08/11/2019	29/11/2023
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	08/11/2019	29/11/2023
							F1903	Device Explantation	A070504	Premature Discharge of Battery	08/11/2019	29/11/2023
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	08/11/2019	29/11/2023
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	08/11/2019	29/11/2023
							F1903	Device Explantation	A070504	Premature Discharge of Battery	08/11/2019	29/11/2023
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	08/11/2019	29/11/2023
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	08/11/2019	29/11/2023
							F1903	Device Explantation	A070504	Premature Discharge of Battery	08/11/2019	29/11/2023
14919	2024	All other reportable incidents	47267	Single-chamber implantable pacemaker, rate-responsive	CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	14/06/2018	29/11/2023
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	14/06/2018	29/11/2023
							F1903	Device Explantation	A070504	Premature Discharge of Battery	14/06/2018	29/11/2023
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	14/06/2018	29/11/2023
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	14/06/2018	29/11/2023
							F1903	Device Explantation	A070504	Premature Discharge of Battery	14/06/2018	29/11/2023
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	14/06/2018	29/11/2023
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	14/06/2018	29/11/2023
							F1903	Device Explantation	A070504	Premature Discharge of Battery	14/06/2018	29/11/2023
14920	2024		47265	Dual-chamber implantable pacemaker, rate-responsive	CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	11/12/2023	11/12/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	11/12/2023	11/12/2023
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	11/12/2023	11/12/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	11/12/2023	11/12/2023
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	11/12/2023	11/12/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	11/12/2023	11/12/2023
14921	2024	All other reportable incidents	47265	Dual-chamber implantable pacemaker, rate-responsive	CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A0708	Power Problem	22/09/2023	22/09/2023
									A0711	Interrogation Problem	22/09/2023	22/09/2023
									A071102	Failure to Interrogate	22/09/2023	22/09/2023
									A071204	No Pacing	22/09/2023	22/09/2023
									A1301	Failure to Read Input Signal	22/09/2023	22/09/2023
									A1302	Failure to Transmit Record	22/09/2023	22/09/2023
							F15	Recognized Device or Procedural Complication	A0708	Power Problem	22/09/2023	22/09/2023
									A0711	Interrogation Problem	22/09/2023	22/09/2023

									A071102	Failure to Interrogate	22/09/2023	22/09/2023
									A071204	No Pacing	22/09/2023	22/09/2023
									A1301	Failure to Read Input Signal	22/09/2023	22/09/2023
									A1302	Failure to Transmit Record	22/09/2023	22/09/2023
							F26	No Health Consequences or Impact	A0708	Power Problem	22/09/2023	22/09/2023
									A0711	Interrogation Problem	22/09/2023	22/09/2023
									A071102	Failure to Interrogate	22/09/2023	22/09/2023
									A071204	No Pacing	22/09/2023	22/09/2023
									A1301	Failure to Read Input Signal	22/09/2023	22/09/2023
									A1302	Failure to Transmit Record	22/09/2023	22/09/2023
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A0708	Power Problem	22/09/2023	22/09/2023
									A0711	Interrogation Problem	22/09/2023	22/09/2023
									A071102	Failure to Interrogate	22/09/2023	22/09/2023
									A071204	No Pacing	22/09/2023	22/09/2023
									A1301	Failure to Read Input Signal	22/09/2023	22/09/2023
									A1302	Failure to Transmit Record	22/09/2023	22/09/2023
							F15	Recognized Device or Procedural Complication	A0708	Power Problem	22/09/2023	22/09/2023
									A0711	Interrogation Problem	22/09/2023	22/09/2023
									A071102	Failure to Interrogate	22/09/2023	22/09/2023
									A071204	No Pacing	22/09/2023	22/09/2023
									A1301	Failure to Read Input Signal	22/09/2023	22/09/2023
									A1302	Failure to Transmit Record	22/09/2023	22/09/2023
							F26	No Health Consequences or Impact	A0708	Power Problem	22/09/2023	22/09/2023
									A0711	Interrogation Problem	22/09/2023	22/09/2023
									A071102	Failure to Interrogate	22/09/2023	22/09/2023
									A071204	No Pacing	22/09/2023	22/09/2023
									A1301	Failure to Read Input Signal	22/09/2023	22/09/2023
									A1302	Failure to Transmit Record	22/09/2023	22/09/2023
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A0708	Power Problem	22/09/2023	22/09/2023
									A0711	Interrogation Problem	22/09/2023	22/09/2023
									A071102	Failure to Interrogate	22/09/2023	22/09/2023
									A071204	No Pacing	22/09/2023	22/09/2023
									A1301	Failure to Read Input Signal	22/09/2023	22/09/2023
									A1302	Failure to Transmit Record	22/09/2023	22/09/2023
							F15	Recognized Device or Procedural Complication	A0708	Power Problem	22/09/2023	22/09/2023
									A0711	Interrogation Problem	22/09/2023	22/09/2023
									A071102	Failure to Interrogate	22/09/2023	22/09/2023
									A071204	No Pacing	22/09/2023	22/09/2023
									A1301	Failure to Read Input Signal	22/09/2023	22/09/2023
									A1302	Failure to Transmit Record	22/09/2023	22/09/2023
							F26	No Health Consequences or Impact	A0708	Power Problem	22/09/2023	22/09/2023
									A0711	Interrogation Problem	22/09/2023	22/09/2023
									A071102	Failure to Interrogate	22/09/2023	22/09/2023
									A071204	No Pacing	22/09/2023	22/09/2023
									A1301	Failure to Read Input Signal	22/09/2023	22/09/2023
									A1302	Failure to Transmit Record	22/09/2023	22/09/2023
14922	2024		47265	Dual-chamber implantable pacemaker, rate-responsive	CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	29/01/2024	29/01/2024
							F1901	Additional Surgery	A24	No Apparent Device Problem	29/01/2024	29/01/2024
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	29/01/2024	29/01/2024
							F1901	Additional Surgery	A24	No Apparent Device Problem	29/01/2024	29/01/2024
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	29/01/2024	29/01/2024
							F1901	Additional Surgery	A24	No Apparent Device Problem	29/01/2024	29/01/2024
14923	2024		47265	Dual-chamber implantable pacemaker, rate-responsive	CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	01/02/2024	01/02/2024
							F1901	Additional Surgery	A24	No Apparent Device Problem	01/02/2024	01/02/2024
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	01/02/2024	01/02/2024
							F1901	Additional Surgery	A24	No Apparent Device Problem	01/02/2024	01/02/2024
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	01/02/2024	01/02/2024
							F1901	Additional Surgery	A24	No Apparent Device Problem	01/02/2024	01/02/2024
14924	2024		47265	Dual-chamber implantable pacemaker, rate-responsive	CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	07/02/2024	07/02/2024
							F1901	Additional Surgery	A24	No Apparent Device Problem	07/02/2024	07/02/2024
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	07/02/2024	07/02/2024
							F1901	Additional Surgery	A24	No Apparent Device Problem	07/02/2024	07/02/2024
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	07/02/2024	07/02/2024
							F1901	Additional Surgery	A24	No Apparent Device Problem	07/02/2024	07/02/2024
14925	2024		47265	Dual-chamber implantable pacemaker, rate-responsive	CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem		
							F1901	Additional Surgery	A24	No Apparent Device Problem		
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem		
							F1901	Additional Surgery	A24	No Apparent Device Problem		
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem		
							F1901	Additional Surgery	A24	No Apparent Device Problem		
14926	2024	All other reportable incidents	35852	Single-chamber implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	23/02/2024	23/02/2024
							F1901	Additional Surgery	A070504	Premature Discharge of Battery	23/02/2024	23/02/2024
							F1903	Device Explantation	A070504	Premature Discharge of Battery	23/02/2024	23/02/2024
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	23/02/2024	23/02/2024
14927	2024	All other reportable incidents	47265	Dual-chamber implantable pacemaker, rate-responsive	CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A070102	High Capture Threshold		
									A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070102	High Capture Threshold		
									A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A070102	High Capture Threshold		
									A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070102	High Capture Threshold		
									A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A070102	High Capture Threshold		
									A0705	Battery Problem		

									A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070102	High Capture Threshold		
									A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
14928	2024	All other reportable incidents	17882	Physiologic-monitoring defibrillation system	CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A150105	Key or Button Unresponsive/not Working		
							F27	Problem Identified During Non-Clinical Procedure	A150105	Key or Button Unresponsive/not Working		
14929	2024		47265	Dual-chamber implantable pacemaker, rate-responsive	CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	02/02/2024	02/02/2024
							F1901	Additional Surgery	A24	No Apparent Device Problem	02/02/2024	02/02/2024
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	02/02/2024	02/02/2024
							F1901	Additional Surgery	A24	No Apparent Device Problem	02/02/2024	02/02/2024
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	02/02/2024	02/02/2024
							F1901	Additional Surgery	A24	No Apparent Device Problem	02/02/2024	02/02/2024
14930	2024		47265	Dual-chamber implantable pacemaker, rate-responsive	CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	08/12/2023	08/12/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	08/12/2023	08/12/2023
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	08/12/2023	08/12/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	08/12/2023	08/12/2023
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	08/12/2023	08/12/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	08/12/2023	08/12/2023
14931	2024	All other reportable incidents	47265	Dual-chamber implantable pacemaker, rate-responsive	CT1685	Cardiac pacemakers	E0717	Dyspnea	A0908	Incorrect, Inadequate or Imprecise Result or Readings	23/02/2024	23/02/2024
							F08	Hospitalization or Prolonged Hospitalization	A0908	Incorrect, Inadequate or Imprecise Result or Readings	23/02/2024	23/02/2024
							F1905	Device Revision or Replacement	A0908	Incorrect, Inadequate or Imprecise Result or Readings	23/02/2024	23/02/2024
					CT257	Implantable cardiac pacemakers	E0717	Dyspnea	A0908	Incorrect, Inadequate or Imprecise Result or Readings	23/02/2024	23/02/2024
							F08	Hospitalization or Prolonged Hospitalization	A0908	Incorrect, Inadequate or Imprecise Result or Readings	23/02/2024	23/02/2024
							F1905	Device Revision or Replacement	A0908	Incorrect, Inadequate or Imprecise Result or Readings	23/02/2024	23/02/2024
					CT2866	Implantable cardiac electrical stimulators	E0717	Dyspnea	A0908	Incorrect, Inadequate or Imprecise Result or Readings	23/02/2024	23/02/2024
							F08	Hospitalization or Prolonged Hospitalization	A0908	Incorrect, Inadequate or Imprecise Result or Readings	23/02/2024	23/02/2024
							F1905	Device Revision or Replacement	A0908	Incorrect, Inadequate or Imprecise Result or Readings	23/02/2024	23/02/2024
14932	2024	All other reportable incidents	47267	Single-chamber implantable pacemaker, rate-responsive	CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
14933	2024	All other reportable incidents	47263	Cardiac resynchronization therapy implantable pacemaker	CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A0908	Incorrect, Inadequate or Imprecise Result or Readings	01/09/2023	01/09/2023
							F08	Hospitalization or Prolonged Hospitalization	A0908	Incorrect, Inadequate or Imprecise Result or Readings	01/09/2023	01/09/2023
							F1905	Device Revision or Replacement	A0908	Incorrect, Inadequate or Imprecise Result or Readings	01/09/2023	01/09/2023
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A0908	Incorrect, Inadequate or Imprecise Result or Readings	01/09/2023	01/09/2023
							F08	Hospitalization or Prolonged Hospitalization	A0908	Incorrect, Inadequate or Imprecise Result or Readings	01/09/2023	01/09/2023
							F1905	Device Revision or Replacement	A0908	Incorrect, Inadequate or Imprecise Result or Readings	01/09/2023	01/09/2023
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A0908	Incorrect, Inadequate or Imprecise Result or Readings	01/09/2023	01/09/2023
							F08	Hospitalization or Prolonged Hospitalization	A0908	Incorrect, Inadequate or Imprecise Result or Readings	01/09/2023	01/09/2023
							F1905	Device Revision or Replacement	A0908	Incorrect, Inadequate or Imprecise Result or Readings	01/09/2023	01/09/2023
14934	2024	All other reportable incidents	47270	Cardiac resynchronization therapy implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A0722	Impedance Problem	08/12/2023	08/12/2023
									A090407	Therapy Delivered to Incorrect Body Area	08/12/2023	08/12/2023
									A1107	Data Problem	08/12/2023	08/12/2023
							F1901	Additional Surgery	A0722	Impedance Problem	08/12/2023	08/12/2023
									A090407	Therapy Delivered to Incorrect Body Area	08/12/2023	08/12/2023
									A1107	Data Problem	08/12/2023	08/12/2023
14935	2024		17882	Physiologic-monitoring defibrillation system	CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
14936	2024	All other reportable incidents	47265	Dual-chamber implantable pacemaker, rate-responsive	CT1685	Cardiac pacemakers	E060101	Asystole	A070101	Failure to Capture		
							F1203	Life Threatening Illness or Injury	A070101	Failure to Capture		
					CT257	Implantable cardiac pacemakers	E060101	Asystole	A070101	Failure to Capture		
							F1203	Life Threatening Illness or Injury	A070101	Failure to Capture		
					CT2866	Implantable cardiac electrical stimulators	E060101	Asystole	A070101	Failure to Capture		
							F1203	Life Threatening Illness or Injury	A070101	Failure to Capture		
14937	2024	All other reportable incidents	63188	Left-ventricular wireless pacing system sensor/transmitter	CT145	Cardiovascular devices	F1903	Device Explantation	A0708	Power Problem		
					CT1577	Wireless monitoring system transmitters	F1903	Device Explantation	A0708	Power Problem		
					CT1684	Cardiac pacemakers and associated devices	F1903	Device Explantation	A0708	Power Problem		
					CT2727	Active implant devices/ancillaries	F1903	Device Explantation	A0708	Power Problem		
					CT2750	Implantable cardiac electrical stimulators and associated devices	F1903	Device Explantation	A0708	Power Problem		
					CT2885	Physiological monitoring/measuring system transmitters/receivers/transceivers	F1903	Device Explantation	A0708	Power Problem		
14938	2024		37265	Dual-chamber implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F1903	Device Explantation	A070504	Premature Discharge of Battery		
14939	2024		47265	Dual-chamber implantable pacemaker, rate-responsive	CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	22/01/2024	22/01/2024
							F1901	Additional Surgery	A24	No Apparent Device Problem	22/01/2024	22/01/2024
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	22/01/2024	22/01/2024
							F1901	Additional Surgery	A24	No Apparent Device Problem	22/01/2024	22/01/2024
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	22/01/2024	22/01/2024
							F1901	Additional Surgery	A24	No Apparent Device Problem	22/01/2024	22/01/2024
14940	2024		47265	Dual-chamber implantable pacemaker, rate-responsive	CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem		
							F1901	Additional Surgery	A24	No Apparent Device Problem		

					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem		
							F1901	Additional Surgery	A24	No Apparent Device Problem		
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem		
							F1901	Additional Surgery	A24	No Apparent Device Problem		
14941	2024	All other reportable incidents	36241	Pacemaker/defibrillator lead	CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A0722	Impedance Problem		
							F08	Hospitalization or Prolonged Hospitalization	A0722	Impedance Problem		
							F1905	Device Revision or Replacement	A0722	Impedance Problem		
14942	2024		47265	Dual-chamber implantable pacemaker, rate-responsive	CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	30/01/2024	30/01/2024
							F1901	Additional Surgery	A24	No Apparent Device Problem	30/01/2024	30/01/2024
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	30/01/2024	30/01/2024
							F1901	Additional Surgery	A24	No Apparent Device Problem	30/01/2024	30/01/2024
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	30/01/2024	30/01/2024
							F1901	Additional Surgery	A24	No Apparent Device Problem	30/01/2024	30/01/2024
14943	2024	All other reportable incidents	47265	Dual-chamber implantable pacemaker, rate-responsive	CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A0508	Noise, Audible		
									A070909	Oversensing		
							F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A0508	Noise, Audible		
									A070909	Oversensing		
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A0508	Noise, Audible		
									A070909	Oversensing		
							F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A0508	Noise, Audible		
									A070909	Oversensing		
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A0508	Noise, Audible		
									A070909	Oversensing		
							F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A0508	Noise, Audible		
									A070909	Oversensing		
14944	2024	All other reportable incidents	47265	Dual-chamber implantable pacemaker, rate-responsive	CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	29/04/2024	29/04/2024
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	29/04/2024	29/04/2024
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	29/04/2024	29/04/2024
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	29/04/2024	29/04/2024
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	29/04/2024	29/04/2024
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	29/04/2024	29/04/2024
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	29/04/2024	29/04/2024
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	29/04/2024	29/04/2024
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	29/04/2024	29/04/2024
14945	2024	All other reportable incidents	36241	Pacemaker/defibrillator lead	CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A0722	Impedance Problem		
							F08	Hospitalization or Prolonged Hospitalization	A0722	Impedance Problem		
							F1905	Device Revision or Replacement	A0722	Impedance Problem		
14946	2024	All other reportable incidents	47270	Cardiac resynchronization therapy implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E011903	Syncope/Fainting	A071201	Failure to Convert Rhythm		
							F23	Unexpected Medical Intervention	A071201	Failure to Convert Rhythm		
14947	2024	Death	47270	Cardiac resynchronization therapy implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E2401	Insufficient Information	A27	Appropriate Device Problem Term/Code not Available		
							F02	Death	A27	Appropriate Device Problem Term/Code not Available		
14948	2024	All other reportable incidents	47265	Dual-chamber implantable pacemaker, rate-responsive	CT1685	Cardiac pacemakers	E1906	Unspecified Infection	A1304	Telemetry Discrepancy	23/02/2024	23/02/2024
							F08	Hospitalization or Prolonged Hospitalization	A1304	Telemetry Discrepancy	23/02/2024	23/02/2024
							F1905	Device Revision or Replacement	A1304	Telemetry Discrepancy	23/02/2024	23/02/2024
					CT257	Implantable cardiac pacemakers	E1906	Unspecified Infection	A1304	Telemetry Discrepancy	23/02/2024	23/02/2024
							F08	Hospitalization or Prolonged Hospitalization	A1304	Telemetry Discrepancy	23/02/2024	23/02/2024
							F1905	Device Revision or Replacement	A1304	Telemetry Discrepancy	23/02/2024	23/02/2024
					CT2866	Implantable cardiac electrical stimulators	E1906	Unspecified Infection	A1304	Telemetry Discrepancy	23/02/2024	23/02/2024
							F08	Hospitalization or Prolonged Hospitalization	A1304	Telemetry Discrepancy	23/02/2024	23/02/2024
							F1905	Device Revision or Replacement	A1304	Telemetry Discrepancy	23/02/2024	23/02/2024
14949	2024	All other reportable incidents	47270	Cardiac resynchronization therapy implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A160601	Premature Elective Replacement Indicator	29/02/2024	29/02/2024
							F1903	Device Explantation	A160601	Premature Elective Replacement Indicator	29/02/2024	29/02/2024
14950	2024		47265	Dual-chamber implantable pacemaker, rate-responsive	CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	06/11/2023	06/11/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	06/11/2023	06/11/2023
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	06/11/2023	06/11/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	06/11/2023	06/11/2023
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	06/11/2023	06/11/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	06/11/2023	06/11/2023
14951	2024	All other reportable incidents	47265	Dual-chamber implantable pacemaker, rate-responsive	CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem	29/09/2023	29/09/2023
									A070504	Premature Discharge of Battery	29/09/2023	29/09/2023
							F08	Hospitalization or Prolonged Hospitalization	A0705	Battery Problem	29/09/2023	29/09/2023
									A070504	Premature Discharge of Battery	29/09/2023	29/09/2023
							F1905	Device Revision or Replacement	A0705	Battery Problem	29/09/2023	29/09/2023
									A070504	Premature Discharge of Battery	29/09/2023	29/09/2023
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem	29/09/2023	29/09/2023
									A070504	Premature Discharge of Battery	29/09/2023	29/09/2023
							F08	Hospitalization or Prolonged Hospitalization	A0705	Battery Problem	29/09/2023	29/09/2023
									A070504	Premature Discharge of Battery	29/09/2023	29/09/2023
							F1905	Device Revision or Replacement	A0705	Battery Problem	29/09/2023	29/09/2023
									A070504	Premature Discharge of Battery	29/09/2023	29/09/2023
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem	29/09/2023	29/09/2023
									A070504	Premature Discharge of Battery	29/09/2023	29/09/2023
							F08	Hospitalization or Prolonged Hospitalization	A0705	Battery Problem	29/09/2023	29/09/2023
									A070504	Premature Discharge of Battery	29/09/2023	29/09/2023
							F1905	Device Revision or Replacement	A0705	Battery Problem	29/09/2023	29/09/2023
									A070504	Premature Discharge of Battery	29/09/2023	29/09/2023
14952	2024	All other reportable incidents	47265	Dual-chamber implantable pacemaker, rate-responsive	CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	29/02/2024	29/02/2024
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	29/02/2024	29/02/2024
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	29/02/2024	29/02/2024
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	29/02/2024	29/02/2024
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	29/02/2024	29/02/2024
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	29/02/2024	29/02/2024
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	29/02/2024	29/02/2024
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	29/02/2024	29/02/2024
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	29/02/2024	29/02/2024

14953	2024	All other reportable incidents	47265	Dual-chamber implantable pacemaker, rate-responsive	CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A0908	Incorrect, Inadequate or Imprecise Result or Readings	01/03/2024	01/03/2024
							F08	Hospitalization or Prolonged Hospitalization	A0908	Incorrect, Inadequate or Imprecise Result or Readings	01/03/2024	01/03/2024
							F1905	Device Revision or Replacement	A0908	Incorrect, Inadequate or Imprecise Result or Readings	01/03/2024	01/03/2024
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A0908	Incorrect, Inadequate or Imprecise Result or Readings	01/03/2024	01/03/2024
							F08	Hospitalization or Prolonged Hospitalization	A0908	Incorrect, Inadequate or Imprecise Result or Readings	01/03/2024	01/03/2024
							F1905	Device Revision or Replacement	A0908	Incorrect, Inadequate or Imprecise Result or Readings	01/03/2024	01/03/2024
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A0908	Incorrect, Inadequate or Imprecise Result or Readings	01/03/2024	01/03/2024
							F08	Hospitalization or Prolonged Hospitalization	A0908	Incorrect, Inadequate or Imprecise Result or Readings	01/03/2024	01/03/2024
							F1905	Device Revision or Replacement	A0908	Incorrect, Inadequate or Imprecise Result or Readings	01/03/2024	01/03/2024
14954	2024	All other reportable incidents	47263	Cardiac resynchronization therapy implantable pacemaker	CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F12	Serious Injury/ Illness/ Impairment	A070504	Premature Discharge of Battery		
							F2307	Reprogramming of Device	A070504	Premature Discharge of Battery		
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F12	Serious Injury/ Illness/ Impairment	A070504	Premature Discharge of Battery		
							F2307	Reprogramming of Device	A070504	Premature Discharge of Battery		
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F12	Serious Injury/ Illness/ Impairment	A070504	Premature Discharge of Battery		
							F2307	Reprogramming of Device	A070504	Premature Discharge of Battery		
14955	2024	All other reportable incidents	47270	Cardiac resynchronization therapy implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
14956	2024		47265	Dual-chamber implantable pacemaker, rate-responsive	CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	09/12/2023	09/12/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	09/12/2023	09/12/2023
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	09/12/2023	09/12/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	09/12/2023	09/12/2023
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	09/12/2023	09/12/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	09/12/2023	09/12/2023
14957	2024		47265	Dual-chamber implantable pacemaker, rate-responsive	CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	09/11/2023	09/11/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	09/11/2023	09/11/2023
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	09/11/2023	09/11/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	09/11/2023	09/11/2023
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	09/11/2023	09/11/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	09/11/2023	09/11/2023
14958	2024	All other reportable incidents	47265	Dual-chamber implantable pacemaker, rate-responsive	CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A05	Mechanical Problem	07/12/2023	07/12/2023
									A0705	Battery Problem	07/12/2023	07/12/2023
									A070504	Premature Discharge of Battery	07/12/2023	07/12/2023
							F08	Hospitalization or Prolonged Hospitalization	A05	Mechanical Problem	07/12/2023	07/12/2023
									A0705	Battery Problem	07/12/2023	07/12/2023
									A070504	Premature Discharge of Battery	07/12/2023	07/12/2023
							F1905	Device Revision or Replacement	A05	Mechanical Problem	07/12/2023	07/12/2023
									A0705	Battery Problem	07/12/2023	07/12/2023
									A070504	Premature Discharge of Battery	07/12/2023	07/12/2023
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A05	Mechanical Problem	07/12/2023	07/12/2023
									A0705	Battery Problem	07/12/2023	07/12/2023
									A070504	Premature Discharge of Battery	07/12/2023	07/12/2023
							F08	Hospitalization or Prolonged Hospitalization	A05	Mechanical Problem	07/12/2023	07/12/2023
									A0705	Battery Problem	07/12/2023	07/12/2023
									A070504	Premature Discharge of Battery	07/12/2023	07/12/2023
							F1905	Device Revision or Replacement	A05	Mechanical Problem	07/12/2023	07/12/2023
									A0705	Battery Problem	07/12/2023	07/12/2023
									A070504	Premature Discharge of Battery	07/12/2023	07/12/2023
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A05	Mechanical Problem	07/12/2023	07/12/2023
									A0705	Battery Problem	07/12/2023	07/12/2023
									A070504	Premature Discharge of Battery	07/12/2023	07/12/2023
							F08	Hospitalization or Prolonged Hospitalization	A05	Mechanical Problem	07/12/2023	07/12/2023
									A0705	Battery Problem	07/12/2023	07/12/2023
									A070504	Premature Discharge of Battery	07/12/2023	07/12/2023
							F1905	Device Revision or Replacement	A05	Mechanical Problem	07/12/2023	07/12/2023
									A0705	Battery Problem	07/12/2023	07/12/2023
									A070504	Premature Discharge of Battery	07/12/2023	07/12/2023
14959	2024	All other reportable incidents	47263	Cardiac resynchronization therapy implantable pacemaker	CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A0908	Incorrect, Inadequate or Imprecise Result or Readings	12/12/2023	12/12/2023
							F08	Hospitalization or Prolonged Hospitalization	A0908	Incorrect, Inadequate or Imprecise Result or Readings	12/12/2023	12/12/2023
							F1905	Device Revision or Replacement	A0908	Incorrect, Inadequate or Imprecise Result or Readings	12/12/2023	12/12/2023
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A0908	Incorrect, Inadequate or Imprecise Result or Readings	12/12/2023	12/12/2023
							F08	Hospitalization or Prolonged Hospitalization	A0908	Incorrect, Inadequate or Imprecise Result or Readings	12/12/2023	12/12/2023
							F1905	Device Revision or Replacement	A0908	Incorrect, Inadequate or Imprecise Result or Readings	12/12/2023	12/12/2023
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A0908	Incorrect, Inadequate or Imprecise Result or Readings	12/12/2023	12/12/2023
							F08	Hospitalization or Prolonged Hospitalization	A0908	Incorrect, Inadequate or Imprecise Result or Readings	12/12/2023	12/12/2023
							F1905	Device Revision or Replacement	A0908	Incorrect, Inadequate or Imprecise Result or Readings	12/12/2023	12/12/2023
14960	2024	All other reportable incidents	47265	Dual-chamber implantable pacemaker, rate-responsive	CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	20/03/2024	20/03/2024
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	20/03/2024	20/03/2024
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	20/03/2024	20/03/2024
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	20/03/2024	20/03/2024
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	20/03/2024	20/03/2024
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	20/03/2024	20/03/2024

					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	20/03/2024	20/03/2024
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	20/03/2024	20/03/2024
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	20/03/2024	20/03/2024
14961	2024	All other reportable incidents	47265	Dual-chamber implantable pacemaker, rate-responsive	CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	24/08/2023	24/08/2023
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	24/08/2023	24/08/2023
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	24/08/2023	24/08/2023
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	24/08/2023	24/08/2023
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	24/08/2023	24/08/2023
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	24/08/2023	24/08/2023
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	24/08/2023	24/08/2023
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	24/08/2023	24/08/2023
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	24/08/2023	24/08/2023
14962	2024	All other reportable incidents	37265	Dual-chamber implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem	01/05/2023	01/05/2023
									A070504	Premature Discharge of Battery	01/05/2023	01/05/2023
							F08	Hospitalization or Prolonged Hospitalization	A0705	Battery Problem	01/05/2023	01/05/2023
									A070504	Premature Discharge of Battery	01/05/2023	01/05/2023
							F1905	Device Revision or Replacement	A0705	Battery Problem	01/05/2023	01/05/2023
									A070504	Premature Discharge of Battery	01/05/2023	01/05/2023
14963	2024	All other reportable incidents	37265	Dual-chamber implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem	18/03/2024	18/03/2024
									A070504	Premature Discharge of Battery	18/03/2024	18/03/2024
									A0908	Incorrect, Inadequate or Imprecise Result or Readings	18/03/2024	18/03/2024
							F08	Hospitalization or Prolonged Hospitalization	A0705	Battery Problem	18/03/2024	18/03/2024
									A070504	Premature Discharge of Battery	18/03/2024	18/03/2024
									A0908	Incorrect, Inadequate or Imprecise Result or Readings	18/03/2024	18/03/2024
							F1905	Device Revision or Replacement	A0705	Battery Problem	18/03/2024	18/03/2024
									A070504	Premature Discharge of Battery	18/03/2024	18/03/2024
									A0908	Incorrect, Inadequate or Imprecise Result or Readings	18/03/2024	18/03/2024
14964	2024	All other reportable incidents	47263	Cardiac resynchronization therapy implantable pacemaker	CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A0908	Incorrect, Inadequate or Imprecise Result or Readings	18/03/2024	18/03/2024
							F08	Hospitalization or Prolonged Hospitalization	A0908	Incorrect, Inadequate or Imprecise Result or Readings	18/03/2024	18/03/2024
							F1905	Device Revision or Replacement	A0908	Incorrect, Inadequate or Imprecise Result or Readings	18/03/2024	18/03/2024
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A0908	Incorrect, Inadequate or Imprecise Result or Readings	18/03/2024	18/03/2024
							F08	Hospitalization or Prolonged Hospitalization	A0908	Incorrect, Inadequate or Imprecise Result or Readings	18/03/2024	18/03/2024
							F1905	Device Revision or Replacement	A0908	Incorrect, Inadequate or Imprecise Result or Readings	18/03/2024	18/03/2024
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A0908	Incorrect, Inadequate or Imprecise Result or Readings	18/03/2024	18/03/2024
							F08	Hospitalization or Prolonged Hospitalization	A0908	Incorrect, Inadequate or Imprecise Result or Readings	18/03/2024	18/03/2024
							F1905	Device Revision or Replacement	A0908	Incorrect, Inadequate or Imprecise Result or Readings	18/03/2024	18/03/2024
14965	2024	All other reportable incidents	47265	Dual-chamber implantable pacemaker, rate-responsive	CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A070909	Oversensing		
									A090801	Signal Artifact/Noise		
							F26	No Health Consequences or Impact	A070909	Oversensing		
									A090801	Signal Artifact/Noise		
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A070909	Oversensing		
									A090801	Signal Artifact/Noise		
							F26	No Health Consequences or Impact	A070909	Oversensing		
									A090801	Signal Artifact/Noise		
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A070909	Oversensing		
									A090801	Signal Artifact/Noise		
							F26	No Health Consequences or Impact	A070909	Oversensing		
									A090801	Signal Artifact/Noise		
14966	2024		47265	Dual-chamber implantable pacemaker, rate-responsive	CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	05/03/2024	05/03/2024
							F1901	Additional Surgery	A24	No Apparent Device Problem	05/03/2024	05/03/2024
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	05/03/2024	05/03/2024
							F1901	Additional Surgery	A24	No Apparent Device Problem	05/03/2024	05/03/2024
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	05/03/2024	05/03/2024
							F1901	Additional Surgery	A24	No Apparent Device Problem	05/03/2024	05/03/2024
14967	2024	All other reportable incidents	17882	Physiologic-monitoring defibrillation system	CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070803	Failure to Power Up		
							F27	Problem Identified During Non-Clinical Procedure	A070803	Failure to Power Up		
14968	2024		17882	Physiologic-monitoring defibrillation system	CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F27	Problem Identified During Non-Clinical Procedure	A070504	Premature Discharge of Battery		
14969	2024		17882	Physiologic-monitoring defibrillation system	CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F27	Problem Identified During Non-Clinical Procedure	A070504	Premature Discharge of Battery		
14970	2024		17882	Physiologic-monitoring defibrillation system	CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F27	Problem Identified During Non-Clinical Procedure	A070504	Premature Discharge of Battery		
14971	2024	All other reportable incidents	47265	Dual-chamber implantable pacemaker, rate-responsive	CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	23/02/2024	23/02/2024
							F1901	Additional Surgery	A24	No Apparent Device Problem	23/02/2024	23/02/2024
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	23/02/2024	23/02/2024
							F1901	Additional Surgery	A24	No Apparent Device Problem	23/02/2024	23/02/2024
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	23/02/2024	23/02/2024
							F1901	Additional Surgery	A24	No Apparent Device Problem	23/02/2024	23/02/2024
14972	2024	All other reportable incidents	47263	Cardiac resynchronization therapy implantable pacemaker	CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A090804	False Positive Result		
							F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A090804	False Positive Result		
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A090804	False Positive Result		
							F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A090804	False Positive Result		
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A090804	False Positive Result		
							F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A090804	False Positive Result		
14973	2024	All other reportable incidents	35852	Single-chamber implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E020201	Anxiety	A070504	Premature Discharge of Battery	22/02/2024	22/02/2024
									A0908	Incorrect, Inadequate or Imprecise Result or Readings	22/02/2024	22/02/2024
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	22/02/2024	22/02/2024

									A0908	Incorrect, Inadequate or Imprecise Result or Readings	22/02/2024	22/02/2024
							F11	Non-Serious Injury/ Illness/ Impairment	A070504	Premature Discharge of Battery	22/02/2024	22/02/2024
									A0908	Incorrect, Inadequate or Imprecise Result or Readings	22/02/2024	22/02/2024
							F1903	Device Explantation	A070504	Premature Discharge of Battery	22/02/2024	22/02/2024
									A0908	Incorrect, Inadequate or Imprecise Result or Readings	22/02/2024	22/02/2024
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	22/02/2024	22/02/2024
									A0908	Incorrect, Inadequate or Imprecise Result or Readings	22/02/2024	22/02/2024
14974	2024	All other reportable incidents	47265	Dual-chamber implantable pacemaker, rate-responsive	CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	30/08/2023	30/08/2023
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	30/08/2023	30/08/2023
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	30/08/2023	30/08/2023
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	30/08/2023	30/08/2023
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	30/08/2023	30/08/2023
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	30/08/2023	30/08/2023
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	30/08/2023	30/08/2023
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	30/08/2023	30/08/2023
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	30/08/2023	30/08/2023
14975	2024	All other reportable incidents	17882	Physiologic-monitoring defibrillation system	CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A1203	Disconnection		
							F27	Problem Identified During Non-Clinical Procedure	A1203	Disconnection		
14976	2024	All other reportable incidents	47265	Dual-chamber implantable pacemaker, rate-responsive	CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem	01/03/2024	01/03/2024
									A070504	Premature Discharge of Battery	01/03/2024	01/03/2024
							F08	Hospitalization or Prolonged Hospitalization	A0705	Battery Problem	01/03/2024	01/03/2024
									A070504	Premature Discharge of Battery	01/03/2024	01/03/2024
							F1905	Device Revision or Replacement	A0705	Battery Problem	01/03/2024	01/03/2024
									A070504	Premature Discharge of Battery	01/03/2024	01/03/2024
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem	01/03/2024	01/03/2024
									A070504	Premature Discharge of Battery	01/03/2024	01/03/2024
							F08	Hospitalization or Prolonged Hospitalization	A0705	Battery Problem	01/03/2024	01/03/2024
									A070504	Premature Discharge of Battery	01/03/2024	01/03/2024
							F1905	Device Revision or Replacement	A0705	Battery Problem	01/03/2024	01/03/2024
									A070504	Premature Discharge of Battery	01/03/2024	01/03/2024
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem	01/03/2024	01/03/2024
									A070504	Premature Discharge of Battery	01/03/2024	01/03/2024
							F08	Hospitalization or Prolonged Hospitalization	A0705	Battery Problem	01/03/2024	01/03/2024
									A070504	Premature Discharge of Battery	01/03/2024	01/03/2024
							F1905	Device Revision or Replacement	A0705	Battery Problem	01/03/2024	01/03/2024
									A070504	Premature Discharge of Battery	01/03/2024	01/03/2024
14977	2024	All other reportable incidents	47205	Cardiac pulse generator programmer	CT1684	Cardiac pacemakers and associated devices	E0601	Arrhythmia	A110303	Program or Algorithm Execution Failure		
							F26	No Health Consequences or Impact	A110303	Program or Algorithm Execution Failure		
14978	2024		37265	Dual-chamber implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A0104	Migration or Expulsion of Device	21/02/2024	21/02/2024
									A0701	Capturing Problem	21/02/2024	21/02/2024
									A070909	Oversensing	21/02/2024	21/02/2024
									A1205	Loose or Intermittent Connection	21/02/2024	21/02/2024
							F1901	Additional Surgery	A0104	Migration or Expulsion of Device	21/02/2024	21/02/2024
									A0701	Capturing Problem	21/02/2024	21/02/2024
									A070909	Oversensing	21/02/2024	21/02/2024
									A1205	Loose or Intermittent Connection	21/02/2024	21/02/2024
14979	2024	All other reportable incidents	47270	Cardiac resynchronization therapy implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E210401	Shock from Patient Lead(s)	A071302	Inappropriate/Inadequate Shock/Stimulation		
							F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A071302	Inappropriate/Inadequate Shock/Stimulation		
14980	2024	All other reportable incidents	37265	Dual-chamber implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A160602	Premature End-of-Life Indicator	23/02/2024	23/02/2024
							F1905	Device Revision or Replacement	A160602	Premature End-of-Life Indicator	23/02/2024	23/02/2024
14981	2024	All other reportable incidents	47265	Dual-chamber implantable pacemaker, rate-responsive	CT1685	Cardiac pacemakers	E0112	Dizziness	A070909	Oversensing	04/02/2024	04/02/2024
									A071204	No Pacing	04/02/2024	04/02/2024
							F1901	Additional Surgery	A070909	Oversensing	04/02/2024	04/02/2024
									A071204	No Pacing	04/02/2024	04/02/2024
					CT257	Implantable cardiac pacemakers	E0112	Dizziness	A070909	Oversensing	04/02/2024	04/02/2024
									A071204	No Pacing	04/02/2024	04/02/2024
							F1901	Additional Surgery	A070909	Oversensing	04/02/2024	04/02/2024
									A071204	No Pacing	04/02/2024	04/02/2024
					CT2866	Implantable cardiac electrical stimulators	E0112	Dizziness	A070909	Oversensing	04/02/2024	04/02/2024
									A071204	No Pacing	04/02/2024	04/02/2024
							F1901	Additional Surgery	A070909	Oversensing	04/02/2024	04/02/2024
									A071204	No Pacing	04/02/2024	04/02/2024
14982	2024	All other reportable incidents	47270	Cardiac resynchronization therapy implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E060109	Tachycardia	A0709	Device Sensing Problem		
									A071301	Failure to Deliver Shock/Stimulation		
							F1203	Life Threatening Illness or Injury	A0709	Device Sensing Problem		
									A071301	Failure to Deliver Shock/Stimulation		
							F2307	Reprogramming of Device	A0709	Device Sensing Problem		
									A071301	Failure to Deliver Shock/Stimulation		
14983	2024	All other reportable incidents	17882	Physiologic-monitoring defibrillation system	CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A090201	Device Displays Incorrect Message		
							F27	Problem Identified During Non-Clinical Procedure	A090201	Device Displays Incorrect Message		
14984	2024		47265	Dual-chamber implantable pacemaker, rate-responsive	CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem		
							F1901	Additional Surgery	A24	No Apparent Device Problem		
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem		
							F1901	Additional Surgery	A24	No Apparent Device Problem		
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem		
							F1901	Additional Surgery	A24	No Apparent Device Problem		
14985	2024	All other reportable incidents	47265	Dual-chamber implantable pacemaker, rate-responsive	CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A0908	Incorrect, Inadequate or Imprecise Result or Readings		
							F26	No Health Consequences or Impact	A0908	Incorrect, Inadequate or Imprecise Result or Readings		
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A0908	Incorrect, Inadequate or Imprecise Result or Readings		
							F26	No Health Consequences or Impact	A0908	Incorrect, Inadequate or Imprecise Result or Readings		
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A0908	Incorrect, Inadequate or Imprecise Result or Readings		

							F26	No Health Consequences or Impact	A0908	Incorrect, Inadequate or Imprecise Result or Readings			
14986	2024	All other reportable incidents	47265	Dual-chamber implantable pacemaker, rate-responsive	CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A1204	Failure to Disconnect	05/03/2024	05/03/2024	
									A150207	Difficult to Remove	05/03/2024	05/03/2024	
							F1901	Additional Surgery	A1204	Failure to Disconnect	05/03/2024	05/03/2024	
									A150207	Difficult to Remove	05/03/2024	05/03/2024	
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A1204	Failure to Disconnect	05/03/2024	05/03/2024	
									A150207	Difficult to Remove	05/03/2024	05/03/2024	
							F1901	Additional Surgery	A1204	Failure to Disconnect	05/03/2024	05/03/2024	
									A150207	Difficult to Remove	05/03/2024	05/03/2024	
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A1204	Failure to Disconnect	05/03/2024	05/03/2024	
									A150207	Difficult to Remove	05/03/2024	05/03/2024	
							F1901	Additional Surgery	A1204	Failure to Disconnect	05/03/2024	05/03/2024	
									A150207	Difficult to Remove	05/03/2024	05/03/2024	
14987	2024	All other reportable incidents	35223	Endocardial/interventricular septal pacing lead	CT1684	Cardiac pacemakers and associated devices	E0601	Arrhythmia	A090801	Signal Artifact/Noise			
							F12	Serious Injury/ Illness/ Impairment	A090801	Signal Artifact/Noise			
							F2307	Reprogramming of Device	A090801	Signal Artifact/Noise			
14988	2024	All other reportable incidents	17882	Physiologic-monitoring defibrillation system	CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070603	Failure to Charge			
							F27	Problem Identified During Non-Clinical Procedure	A070603	Failure to Charge			
14989	2024	All other reportable incidents	47265	Dual-chamber implantable pacemaker, rate-responsive	CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	04/12/2023	04/12/2023	
							F1901	Additional Surgery	A24	No Apparent Device Problem	04/12/2023	04/12/2023	
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	04/12/2023	04/12/2023	
							F1901	Additional Surgery	A24	No Apparent Device Problem	04/12/2023	04/12/2023	
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	04/12/2023	04/12/2023	
							F1901	Additional Surgery	A24	No Apparent Device Problem	04/12/2023	04/12/2023	
14990	2024	All other reportable incidents	35852	Single-chamber implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem	14/03/2024	14/03/2024	
									A070504	Premature Discharge of Battery	14/03/2024	14/03/2024	
									A0908	Incorrect, Inadequate or Imprecise Result or Readings	14/03/2024	14/03/2024	
							F08	Hospitalization or Prolonged Hospitalization	A0705	Battery Problem	14/03/2024	14/03/2024	
									A070504	Premature Discharge of Battery	14/03/2024	14/03/2024	
									A0908	Incorrect, Inadequate or Imprecise Result or Readings	14/03/2024	14/03/2024	
							F1905	Device Revision or Replacement	A0705	Battery Problem	14/03/2024	14/03/2024	
									A070504	Premature Discharge of Battery	14/03/2024	14/03/2024	
									A0908	Incorrect, Inadequate or Imprecise Result or Readings	14/03/2024	14/03/2024	
14991	2024	All other reportable incidents	36241	Pacemaker/defibrillator lead	CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070909	Oversensing			
									A090801	Signal Artifact/Noise			
							F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A070909	Oversensing			
									A090801	Signal Artifact/Noise			
14992	2024	All other reportable incidents	47265	Dual-chamber implantable pacemaker, rate-responsive	CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A070909	Oversensing	01/04/2024	01/04/2024	
									A0712	Pacing Problem	01/04/2024	01/04/2024	
									A090801	Signal Artifact/Noise	01/04/2024	01/04/2024	
							F08	Hospitalization or Prolonged Hospitalization	A070909	Oversensing	01/04/2024	01/04/2024	
									A0712	Pacing Problem	01/04/2024	01/04/2024	
									A090801	Signal Artifact/Noise	01/04/2024	01/04/2024	
							F1001	Absence of Treatment	A070909	Oversensing	01/04/2024	01/04/2024	
									A0712	Pacing Problem	01/04/2024	01/04/2024	
									A090801	Signal Artifact/Noise	01/04/2024	01/04/2024	
							F1905	Device Revision or Replacement	A070909	Oversensing	01/04/2024	01/04/2024	
									A0712	Pacing Problem	01/04/2024	01/04/2024	
									A090801	Signal Artifact/Noise	01/04/2024	01/04/2024	
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A070909	Oversensing	01/04/2024	01/04/2024	
									A0712	Pacing Problem	01/04/2024	01/04/2024	
									A090801	Signal Artifact/Noise	01/04/2024	01/04/2024	
							F08	Hospitalization or Prolonged Hospitalization	A070909	Oversensing	01/04/2024	01/04/2024	
									A0712	Pacing Problem	01/04/2024	01/04/2024	
									A090801	Signal Artifact/Noise	01/04/2024	01/04/2024	
							F1001	Absence of Treatment	A070909	Oversensing	01/04/2024	01/04/2024	
									A0712	Pacing Problem	01/04/2024	01/04/2024	
									A090801	Signal Artifact/Noise	01/04/2024	01/04/2024	
							F1905	Device Revision or Replacement	A070909	Oversensing	01/04/2024	01/04/2024	
									A0712	Pacing Problem	01/04/2024	01/04/2024	
									A090801	Signal Artifact/Noise	01/04/2024	01/04/2024	
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A070909	Oversensing	01/04/2024	01/04/2024	
									A0712	Pacing Problem	01/04/2024	01/04/2024	
									A090801	Signal Artifact/Noise	01/04/2024	01/04/2024	
							F08	Hospitalization or Prolonged Hospitalization	A070909	Oversensing	01/04/2024	01/04/2024	
									A0712	Pacing Problem	01/04/2024	01/04/2024	
									A090801	Signal Artifact/Noise	01/04/2024	01/04/2024	
							F1001	Absence of Treatment	A070909	Oversensing	01/04/2024	01/04/2024	
									A0712	Pacing Problem	01/04/2024	01/04/2024	
									A090801	Signal Artifact/Noise	01/04/2024	01/04/2024	
							F1905	Device Revision or Replacement	A070909	Oversensing	01/04/2024	01/04/2024	
									A0712	Pacing Problem	01/04/2024	01/04/2024	
									A090801	Signal Artifact/Noise	01/04/2024	01/04/2024	
14993	2024	All other reportable incidents			CT1684	Cardiac pacemakers and associated devices	E2104	Electric Shock	A071302	Inappropriate/Inadequate Shock/Stimulation	29/02/2024	29/02/2024	
							F08	Hospitalization or Prolonged Hospitalization	A071302	Inappropriate/Inadequate Shock/Stimulation	29/02/2024	29/02/2024	
							F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A071302	Inappropriate/Inadequate Shock/Stimulation	29/02/2024	29/02/2024	
14994	2024	All other reportable incidents	37265	Dual-chamber implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E210401	Shock from Patient Lead(s)	A070909	Oversensing			
									A071302	Inappropriate/Inadequate Shock/Stimulation			
							F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A070909	Oversensing			
									A071302	Inappropriate/Inadequate Shock/Stimulation			
14995	2024	All other reportable incidents	37265	Dual-chamber implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A070909	Oversensing			
									A0712	Pacing Problem			
							F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A070909	Oversensing			
									A0712	Pacing Problem			
14996	2024	All other reportable incidents	47265	Dual-chamber implantable pacemaker, rate-	CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A05	Mechanical Problem			

				responsive			F23	Unexpected Medical Intervention	A05	Mechanical Problem		
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A05	Mechanical Problem		
							F23	Unexpected Medical Intervention	A05	Mechanical Problem		
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A05	Mechanical Problem		
							F23	Unexpected Medical Intervention	A05	Mechanical Problem		
14997	2024	All other reportable incidents	47265	Dual-chamber implantable pacemaker, rate-responsive	CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	06/03/2024	06/03/2024
							F1901	Additional Surgery	A24	No Apparent Device Problem	06/03/2024	06/03/2024
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	06/03/2024	06/03/2024
							F1901	Additional Surgery	A24	No Apparent Device Problem	06/03/2024	06/03/2024
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	06/03/2024	06/03/2024
							F1901	Additional Surgery	A24	No Apparent Device Problem	06/03/2024	06/03/2024
14998	2024	All other reportable incidents	47267	Single-chamber implantable pacemaker, rate-responsive			E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem	23/01/2025	23/01/2025
									A070504	Premature Discharge of Battery	23/01/2025	23/01/2025
							F08	Hospitalization or Prolonged Hospitalization	A0705	Battery Problem	23/01/2025	23/01/2025
									A070504	Premature Discharge of Battery	23/01/2025	23/01/2025
							F1905	Device Revision or Replacement	A0705	Battery Problem	23/01/2025	23/01/2025
									A070504	Premature Discharge of Battery	23/01/2025	23/01/2025
14999	2024	All other reportable incidents	47265	Dual-chamber implantable pacemaker, rate-responsive	CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	06/03/2024	06/03/2024
							F1901	Additional Surgery	A24	No Apparent Device Problem	06/03/2024	06/03/2024
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	06/03/2024	06/03/2024
							F1901	Additional Surgery	A24	No Apparent Device Problem	06/03/2024	06/03/2024
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	06/03/2024	06/03/2024
							F1901	Additional Surgery	A24	No Apparent Device Problem	06/03/2024	06/03/2024
15000	2024	All other reportable incidents	47263	Cardiac resynchronization therapy implantable pacemaker	CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A070909	Oversensing		
									A0712	Pacing Problem		
									A0908	Incorrect, Inadequate or Imprecise Result or Readings		
							F26	No Health Consequences or Impact	A070909	Oversensing		
									A0712	Pacing Problem		
									A0908	Incorrect, Inadequate or Imprecise Result or Readings		
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A070909	Oversensing		
									A0712	Pacing Problem		
									A0908	Incorrect, Inadequate or Imprecise Result or Readings		
							F26	No Health Consequences or Impact	A070909	Oversensing		
									A0712	Pacing Problem		
									A0908	Incorrect, Inadequate or Imprecise Result or Readings		
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A070909	Oversensing		
									A0712	Pacing Problem		
									A0908	Incorrect, Inadequate or Imprecise Result or Readings		
							F26	No Health Consequences or Impact	A070909	Oversensing		
									A0712	Pacing Problem		
									A0908	Incorrect, Inadequate or Imprecise Result or Readings		
15001	2024	All other reportable incidents	47265	Dual-chamber implantable pacemaker, rate-responsive	CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem		
							F1901	Additional Surgery	A24	No Apparent Device Problem		
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem		
							F1901	Additional Surgery	A24	No Apparent Device Problem		
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem		
							F1901	Additional Surgery	A24	No Apparent Device Problem		
15002	2024		47265	Dual-chamber implantable pacemaker, rate-responsive	CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	11/03/2024	11/03/2024
							F1901	Additional Surgery	A24	No Apparent Device Problem	11/03/2024	11/03/2024
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	11/03/2024	11/03/2024
							F1901	Additional Surgery	A24	No Apparent Device Problem	11/03/2024	11/03/2024
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	11/03/2024	11/03/2024
							F1901	Additional Surgery	A24	No Apparent Device Problem	11/03/2024	11/03/2024
15003	2024	All other reportable incidents	37265	Dual-chamber implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E0601	Arrhythmia	A071201	Failure to Convert Rhythm		
							F12	Serious Injury/ Illness/ Impairment	A071201	Failure to Convert Rhythm		
15004	2024	All other reportable incidents	47270	Cardiac resynchronization therapy implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E060109	Tachycardia	A0709	Device Sensing Problem		
									A071301	Failure to Deliver Shock/Stimulation		
							F1203	Life Threatening Illness or Injury	A0709	Device Sensing Problem		
									A071301	Failure to Deliver Shock/Stimulation		
							F2307	Reprogramming of Device	A0709	Device Sensing Problem		
									A071301	Failure to Deliver Shock/Stimulation		
15005	2024		60789	Intracardiac pacemaker	CT1685	Cardiac pacemakers	E0506	Hemorrhage/Blood Loss/Bleeding	A0512	Unintended Movement		
							E0604	Cardiac Perforation	A0512	Unintended Movement		
							E0605	Cardiac Tamponade	A0512	Unintended Movement		
							E0619	Pericardial Effusion	A0512	Unintended Movement		
							F15	Recognized Device or Procedural Complication	A0512	Unintended Movement		
							F19	Surgical Intervention	A0512	Unintended Movement		
					CT257	Implantable cardiac pacemakers	E0506	Hemorrhage/Blood Loss/Bleeding	A0512	Unintended Movement		
							E0604	Cardiac Perforation	A0512	Unintended Movement		
							E0605	Cardiac Tamponade	A0512	Unintended Movement		
							E0619	Pericardial Effusion	A0512	Unintended Movement		
							F15	Recognized Device or Procedural Complication	A0512	Unintended Movement		
							F19	Surgical Intervention	A0512	Unintended Movement		
					CT2818	Catheterization/Cannulation/Fluid path devices	E0506	Hemorrhage/Blood Loss/Bleeding	A0512	Unintended Movement		
							E0604	Cardiac Perforation	A0512	Unintended Movement		
							E0605	Cardiac Tamponade	A0512	Unintended Movement		
							E0619	Pericardial Effusion	A0512	Unintended Movement		
							F15	Recognized Device or Procedural Complication	A0512	Unintended Movement		
							F19	Surgical Intervention	A0512	Unintended Movement		
					CT2866	Implantable cardiac electrical stimulators	E0506	Hemorrhage/Blood Loss/Bleeding	A0512	Unintended Movement		
							E0604	Cardiac Perforation	A0512	Unintended Movement		
							E0605	Cardiac Tamponade	A0512	Unintended Movement		
							E0619	Pericardial Effusion	A0512	Unintended Movement		

							F15	Recognized Device or Procedural Complication	A0512	Unintended Movement		
							F19	Surgical Intervention	A0512	Unintended Movement		
15006	2024		47267	Single-chamber implantable pacemaker, rate-responsive	CT1685	Cardiac pacemakers	E0717	Dyspnea	A071205	Pacemaker Found in Back-Up Mode		
							F23	Unexpected Medical Intervention	A071205	Pacemaker Found in Back-Up Mode		
					CT257	Implantable cardiac pacemakers	E0717	Dyspnea	A071205	Pacemaker Found in Back-Up Mode		
							F23	Unexpected Medical Intervention	A071205	Pacemaker Found in Back-Up Mode		
					CT2866	Implantable cardiac electrical stimulators	E0717	Dyspnea	A071205	Pacemaker Found in Back-Up Mode		
							F23	Unexpected Medical Intervention	A071205	Pacemaker Found in Back-Up Mode		
15007	2024	All other reportable incidents	47265	Dual-chamber implantable pacemaker, rate-responsive	CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem		
							F26	No Health Consequences or Impact	A0705	Battery Problem		
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem		
							F26	No Health Consequences or Impact	A0705	Battery Problem		
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem		
							F26	No Health Consequences or Impact	A0705	Battery Problem		
15008	2024	All other reportable incidents	47265	Dual-chamber implantable pacemaker, rate-responsive	CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	17/11/2023	17/11/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	17/11/2023	17/11/2023
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	17/11/2023	17/11/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	17/11/2023	17/11/2023
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	17/11/2023	17/11/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	17/11/2023	17/11/2023
15009	2024	All other reportable incidents	35852	Single-chamber implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A0508	Noise, Audible		
									A070909	Oversensing		
									A072201	High Impedance		
							F26	No Health Consequences or Impact	A0508	Noise, Audible		
									A070909	Oversensing		
									A072201	High Impedance		
15010	2024	All other reportable incidents	17882	Physiologic-monitoring defibrillation system	CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A090101	Inappropriate Audible Prompt/Feedback		
							F26	No Health Consequences or Impact	A090101	Inappropriate Audible Prompt/Feedback		
15011	2024		47265	Dual-chamber implantable pacemaker, rate-responsive	CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	11/03/2024	11/03/2024
							F1901	Additional Surgery	A24	No Apparent Device Problem	11/03/2024	11/03/2024
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	11/03/2024	11/03/2024
							F1901	Additional Surgery	A24	No Apparent Device Problem	11/03/2024	11/03/2024
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	11/03/2024	11/03/2024
							F1901	Additional Surgery	A24	No Apparent Device Problem	11/03/2024	11/03/2024
15012	2024	All other reportable incidents	47265	Dual-chamber implantable pacemaker, rate-responsive	CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	07/03/2024	07/03/2024
							F1901	Additional Surgery	A24	No Apparent Device Problem	07/03/2024	07/03/2024
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	07/03/2024	07/03/2024
							F1901	Additional Surgery	A24	No Apparent Device Problem	07/03/2024	07/03/2024
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	07/03/2024	07/03/2024
							F1901	Additional Surgery	A24	No Apparent Device Problem	07/03/2024	07/03/2024
15013	2024	Unanticipated serious deterioration in state of health	35224	External pacemaker, epicardial pacing	CT1685	Cardiac pacemakers	E0602	Cardiac Arrest	A01	Patient Device Interaction Problem		
									A050502	Misfire		
									A0709	Device Sensing Problem		
							F1203	Life Threatening Illness or Injury	A01	Patient Device Interaction Problem		
									A050502	Misfire		
									A0709	Device Sensing Problem		
15014	2024		18111	Pacing lead extraction kit, reusable	CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070102	High Capture Threshold		
									A072201	High Impedance		
							F27	Problem Identified During Non-Clinical Procedure	A070102	High Capture Threshold		
									A072201	High Impedance		
15015	2024	All other reportable incidents	35852	Single-chamber implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E210401	Shock from Patient Lead(s)	A071302	Inappropriate/Inadequate Shock/Stimulation		
							F08	Hospitalization or Prolonged Hospitalization	A071302	Inappropriate/Inadequate Shock/Stimulation		
							F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A071302	Inappropriate/Inadequate Shock/Stimulation		
							F2203	Imaging Required	A071302	Inappropriate/Inadequate Shock/Stimulation		
15016	2024	All other reportable incidents	35852	Single-chamber implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E210401	Shock from Patient Lead(s)	A071302	Inappropriate/Inadequate Shock/Stimulation		
							F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A071302	Inappropriate/Inadequate Shock/Stimulation		
15017	2024	All other reportable incidents	35852	Single-chamber implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A072201	High Impedance		
									A2304	Off-Label Use		
							F26	No Health Consequences or Impact	A072201	High Impedance		
									A2304	Off-Label Use		
15018	2024	All other reportable incidents	47267	Single-chamber implantable pacemaker, rate-responsive	CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	15/03/2024	15/03/2024
							F1901	Additional Surgery	A070504	Premature Discharge of Battery	15/03/2024	15/03/2024
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	15/03/2024	15/03/2024
							F1901	Additional Surgery	A070504	Premature Discharge of Battery	15/03/2024	15/03/2024
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	15/03/2024	15/03/2024
							F1901	Additional Surgery	A070504	Premature Discharge of Battery	15/03/2024	15/03/2024
15019	2024	All other reportable incidents			CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070910	Undersensing	29/08/2023	29/08/2023
									A0722	Impedance Problem	29/08/2023	29/08/2023
							F08	Hospitalization or Prolonged Hospitalization	A070910	Undersensing	29/08/2023	29/08/2023
									A0722	Impedance Problem	29/08/2023	29/08/2023
							F1905	Device Revision or Replacement	A070910	Undersensing	29/08/2023	29/08/2023
									A0722	Impedance Problem	29/08/2023	29/08/2023
15020	2024	All other reportable incidents	47265	Dual-chamber implantable pacemaker, rate-responsive	CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	01/11/2023	01/11/2023
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	01/11/2023	01/11/2023
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	01/11/2023	01/11/2023
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	01/11/2023	01/11/2023
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	01/11/2023	01/11/2023
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	01/11/2023	01/11/2023
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	01/11/2023	01/11/2023
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	01/11/2023	01/11/2023
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	01/11/2023	01/11/2023
15021	2024	All other reportable incidents	17882	Physiologic-monitoring defibrillation system	CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070803	Failure to Power Up		
							F27	Problem Identified During Non-Clinical Procedure	A070803	Failure to Power Up		
15022	2024	All other reportable incidents	47265	Dual-chamber implantable pacemaker, rate-responsive	CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem	28/09/2023	28/09/2023
									A070504	Premature Discharge of Battery	28/09/2023	28/09/2023
							F08	Hospitalization or Prolonged Hospitalization	A0705	Battery Problem	28/09/2023	28/09/2023
									A070504	Premature Discharge of Battery	28/09/2023	28/09/2023
							F1905	Device Revision or Replacement	A0705	Battery Problem	28/09/2023	28/09/2023

									A070504	Premature Discharge of Battery	28/09/2023	28/09/2023
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem	28/09/2023	28/09/2023
									A070504	Premature Discharge of Battery	28/09/2023	28/09/2023
							F08	Hospitalization or Prolonged Hospitalization	A0705	Battery Problem	28/09/2023	28/09/2023
									A070504	Premature Discharge of Battery	28/09/2023	28/09/2023
							F1905	Device Revision or Replacement	A0705	Battery Problem	28/09/2023	28/09/2023
									A070504	Premature Discharge of Battery	28/09/2023	28/09/2023
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem	28/09/2023	28/09/2023
									A070504	Premature Discharge of Battery	28/09/2023	28/09/2023
							F08	Hospitalization or Prolonged Hospitalization	A0705	Battery Problem	28/09/2023	28/09/2023
									A070504	Premature Discharge of Battery	28/09/2023	28/09/2023
							F1905	Device Revision or Replacement	A0705	Battery Problem	28/09/2023	28/09/2023
									A070504	Premature Discharge of Battery	28/09/2023	28/09/2023
15023	2024		47265	Dual-chamber implantable pacemaker, rate-responsive	CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	08/02/2024	08/02/2024
							F1901	Additional Surgery	A24	No Apparent Device Problem	08/02/2024	08/02/2024
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	08/02/2024	08/02/2024
							F1901	Additional Surgery	A24	No Apparent Device Problem	08/02/2024	08/02/2024
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	08/02/2024	08/02/2024
							F1901	Additional Surgery	A24	No Apparent Device Problem	08/02/2024	08/02/2024
15024	2024	All other reportable incidents	17882	Physiologic-monitoring defibrillation system	CT1684	Cardiac pacemakers and associated devices	E0601	Arrhythmia	A0713	Defibrillation/Stimulation Problem		
							F05	Delay to Treatment/ Therapy	A0713	Defibrillation/Stimulation Problem		
15025	2024	All other reportable incidents	47265	Dual-chamber implantable pacemaker, rate-responsive	CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem		
							F1901	Additional Surgery	A24	No Apparent Device Problem		
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem		
							F1901	Additional Surgery	A24	No Apparent Device Problem		
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem		
							F1901	Additional Surgery	A24	No Apparent Device Problem		
15026	2024	All other reportable incidents	37265	Dual-chamber implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A0401	Break		
									A150201	Positioning Failure		
									A23	Use of Device Problem		
							F1908	Prolonged Surgery	A0401	Break		
									A150201	Positioning Failure		
									A23	Use of Device Problem		
15027	2024	All other reportable incidents	47270	Cardiac resynchronization therapy implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A0722	Impedance Problem		
									A1301	Failure to Read Input Signal		
							F26	No Health Consequences or Impact	A0722	Impedance Problem		
									A1301	Failure to Read Input Signal		
15028	2024	All other reportable incidents	47270	Cardiac resynchronization therapy implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A072201	High Impedance		
							F26	No Health Consequences or Impact	A072201	High Impedance		
15029	2024	All other reportable incidents	17882	Physiologic-monitoring defibrillation system	CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A0710	Failure to Conduct		
							F27	Problem Identified During Non-Clinical Procedure	A0710	Failure to Conduct		
15030	2024	All other reportable incidents	17882	Physiologic-monitoring defibrillation system	CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A0719	Unexpected Shutdown		
							F27	Problem Identified During Non-Clinical Procedure	A0719	Unexpected Shutdown		
15031	2024	All other reportable incidents	37265	Dual-chamber implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E011903	Syncope/Fainting	A070909	Oversensing		
									A071201	Failure to Convert Rhythm		
									A071302	Inappropriate/Inadequate Shock/Stimulation		
							E210401	Shock from Patient Lead(s)	A070909	Oversensing		
									A071201	Failure to Convert Rhythm		
									A071302	Inappropriate/Inadequate Shock/Stimulation		
							F08	Hospitalization or Prolonged Hospitalization	A070909	Oversensing		
									A071201	Failure to Convert Rhythm		
									A071302	Inappropriate/Inadequate Shock/Stimulation		
							F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A070909	Oversensing		
									A071201	Failure to Convert Rhythm		
									A071302	Inappropriate/Inadequate Shock/Stimulation		
15032	2024	All other reportable incidents	60789	Intracardiac pacemaker	CT1685	Cardiac pacemakers	E0604	Cardiac Perforation	A24	No Apparent Device Problem		
							E0605	Cardiac Tamponade	A24	No Apparent Device Problem		
							E0619	Pericardial Effusion	A24	No Apparent Device Problem		
							F1203	Life Threatening Illness or Injury	A24	No Apparent Device Problem		
							F1901	Additional Surgery	A24	No Apparent Device Problem		
					CT257	Implantable cardiac pacemakers	E0604	Cardiac Perforation	A24	No Apparent Device Problem		
							E0605	Cardiac Tamponade	A24	No Apparent Device Problem		
							E0619	Pericardial Effusion	A24	No Apparent Device Problem		
							F1203	Life Threatening Illness or Injury	A24	No Apparent Device Problem		
							F1901	Additional Surgery	A24	No Apparent Device Problem		
					CT2818	Catheterization/Cannulation/Fluid path devices	E0604	Cardiac Perforation	A24	No Apparent Device Problem		
							E0605	Cardiac Tamponade	A24	No Apparent Device Problem		
							E0619	Pericardial Effusion	A24	No Apparent Device Problem		
							F1203	Life Threatening Illness or Injury	A24	No Apparent Device Problem		
							F1901	Additional Surgery	A24	No Apparent Device Problem		
					CT2866	Implantable cardiac electrical stimulators	E0604	Cardiac Perforation	A24	No Apparent Device Problem		
							E0605	Cardiac Tamponade	A24	No Apparent Device Problem		
							E0619	Pericardial Effusion	A24	No Apparent Device Problem		
							F1203	Life Threatening Illness or Injury	A24	No Apparent Device Problem		
							F1901	Additional Surgery	A24	No Apparent Device Problem		
15033	2024		47265	Dual-chamber implantable pacemaker, rate-responsive	CT1685	Cardiac pacemakers	E020201	Anxiety	A0203	Defective Device		
									A070909	Oversensing		
									A071205	Pacemaker Found in Back-Up Mode		
							E020203	Distress	A0203	Defective Device		
									A070909	Oversensing		
									A071205	Pacemaker Found in Back-Up Mode		
							F01	Change in Therapeutic Response	A0203	Defective Device		
									A070909	Oversensing		
									A071205	Pacemaker Found in Back-Up Mode		
							F1905	Device Revision or Replacement	A0203	Defective Device		
									A070909	Oversensing		
									A071205	Pacemaker Found in Back-Up Mode		
					CT257	Implantable cardiac pacemakers	E020201	Anxiety	A0203	Defective Device		

									A070909	Oversensing		
									A071205	Pacemaker Found in Back-Up Mode		
							E020203	Distress	A0203	Defective Device		
									A070909	Oversensing		
									A071205	Pacemaker Found in Back-Up Mode		
							F01	Change in Therapeutic Response	A0203	Defective Device		
									A070909	Oversensing		
									A071205	Pacemaker Found in Back-Up Mode		
							F1905	Device Revision or Replacement	A0203	Defective Device		
									A070909	Oversensing		
									A071205	Pacemaker Found in Back-Up Mode		
					CT2866	Implantable cardiac electrical stimulators	E020201	Anxiety	A0203	Defective Device		
									A070909	Oversensing		
									A071205	Pacemaker Found in Back-Up Mode		
							E020203	Distress	A0203	Defective Device		
									A070909	Oversensing		
									A071205	Pacemaker Found in Back-Up Mode		
							F01	Change in Therapeutic Response	A0203	Defective Device		
									A070909	Oversensing		
									A071205	Pacemaker Found in Back-Up Mode		
							F1905	Device Revision or Replacement	A0203	Defective Device		
									A070909	Oversensing		
									A071205	Pacemaker Found in Back-Up Mode		
15034	2024	All other reportable incidents	35852	Single-chamber implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E060102	Atrial Fibrillation	A0712	Pacing Problem		
							F12	Serious Injury/ Illness/ Impairment	A0712	Pacing Problem		
							F23	Unexpected Medical Intervention	A0712	Pacing Problem		
15035	2024	All other reportable incidents	17882	Physiologic-monitoring defibrillation system	CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A0719	Unexpected Shutdown		
							F27	Problem Identified During Non-Clinical Procedure	A0719	Unexpected Shutdown		
15036	2024	All other reportable incidents	47263	Cardiac resynchronization therapy implantable pacemaker	CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A0908	Incorrect, Inadequate or Imprecise Result or Readings	21/03/2024	21/03/2024
							F08	Hospitalization or Prolonged Hospitalization	A0908	Incorrect, Inadequate or Imprecise Result or Readings	21/03/2024	21/03/2024
							F1905	Device Revision or Replacement	A0908	Incorrect, Inadequate or Imprecise Result or Readings	21/03/2024	21/03/2024
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A0908	Incorrect, Inadequate or Imprecise Result or Readings	21/03/2024	21/03/2024
							F08	Hospitalization or Prolonged Hospitalization	A0908	Incorrect, Inadequate or Imprecise Result or Readings	21/03/2024	21/03/2024
							F1905	Device Revision or Replacement	A0908	Incorrect, Inadequate or Imprecise Result or Readings	21/03/2024	21/03/2024
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A0908	Incorrect, Inadequate or Imprecise Result or Readings	21/03/2024	21/03/2024
							F08	Hospitalization or Prolonged Hospitalization	A0908	Incorrect, Inadequate or Imprecise Result or Readings	21/03/2024	21/03/2024
							F1905	Device Revision or Replacement	A0908	Incorrect, Inadequate or Imprecise Result or Readings	21/03/2024	21/03/2024
15037	2024	All other reportable incidents	47265	Dual-chamber implantable pacemaker, rate-responsive	CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A072201	High Impedance		
							F26	No Health Consequences or Impact	A072201	High Impedance		
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A072201	High Impedance		
							F26	No Health Consequences or Impact	A072201	High Impedance		
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A072201	High Impedance		
							F26	No Health Consequences or Impact	A072201	High Impedance		
15038	2024	All other reportable incidents	37265	Dual-chamber implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A160602	Premature End-of-Life Indicator		
							F1903	Device Explantation	A160602	Premature End-of-Life Indicator		
15039	2024	All other reportable incidents	63188	Left-ventricular wireless pacing system sensor/transmitter	CT145	Cardiovascular devices			A0708	Power Problem		
					CT1577	Wireless monitoring system transmitters			A0708	Power Problem		
					CT1684	Cardiac pacemakers and associated devices			A0708	Power Problem		
					CT2727	Active implant devices/ancillaries			A0708	Power Problem		
					CT2750	Implantable cardiac electrical stimulators and associated devices			A0708	Power Problem		
					CT2885	Physiological monitoring/measuring system transmitters/receivers/transceivers			A0708	Power Problem		
15040	2024	All other reportable incidents	47265	Dual-chamber implantable pacemaker, rate-responsive	CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	11/03/2024	11/03/2024
							F1901	Additional Surgery	A24	No Apparent Device Problem	11/03/2024	11/03/2024
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	11/03/2024	11/03/2024
							F1901	Additional Surgery	A24	No Apparent Device Problem	11/03/2024	11/03/2024
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	11/03/2024	11/03/2024
							F1901	Additional Surgery	A24	No Apparent Device Problem	11/03/2024	11/03/2024
15041	2024	All other reportable incidents	17882	Physiologic-monitoring defibrillation system	CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070803	Failure to Power Up		
							F27	Problem Identified During Non-Clinical Procedure	A070803	Failure to Power Up		
15042	2024		47265	Dual-chamber implantable pacemaker, rate-responsive	CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	15/03/2024	15/03/2024
							F1901	Additional Surgery	A24	No Apparent Device Problem	15/03/2024	15/03/2024
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	15/03/2024	15/03/2024
							F1901	Additional Surgery	A24	No Apparent Device Problem	15/03/2024	15/03/2024
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	15/03/2024	15/03/2024
							F1901	Additional Surgery	A24	No Apparent Device Problem	15/03/2024	15/03/2024
15043	2024	All other reportable incidents	47265	Dual-chamber implantable pacemaker, rate-responsive	CT1685	Cardiac pacemakers	E0601	Arrhythmia	A071102	Failure to Interrogate	21/03/2024	21/03/2024
									A071204	No Pacing	21/03/2024	21/03/2024
							E2312	Fatigue	A071102	Failure to Interrogate	21/03/2024	21/03/2024
									A071204	No Pacing	21/03/2024	21/03/2024
							F1901	Additional Surgery	A071102	Failure to Interrogate	21/03/2024	21/03/2024
									A071204	No Pacing	21/03/2024	21/03/2024
					CT257	Implantable cardiac pacemakers	E0601	Arrhythmia	A071102	Failure to Interrogate	21/03/2024	21/03/2024
									A071204	No Pacing	21/03/2024	21/03/2024
							E2312	Fatigue	A071102	Failure to Interrogate	21/03/2024	21/03/2024
									A071204	No Pacing	21/03/2024	21/03/2024
							F1901	Additional Surgery	A071102	Failure to Interrogate	21/03/2024	21/03/2024
									A071204	No Pacing	21/03/2024	21/03/2024
					CT2866	Implantable cardiac electrical stimulators	E0601	Arrhythmia	A071102	Failure to Interrogate	21/03/2024	21/03/2024
									A071204	No Pacing	21/03/2024	21/03/2024

							E2312	Fatigue	A071102	Failure to Interrogate	21/03/2024	21/03/2024
									A071204	No Pacing	21/03/2024	21/03/2024
							F1901	Additional Surgery	A071102	Failure to Interrogate	21/03/2024	21/03/2024
									A071204	No Pacing	21/03/2024	21/03/2024
15044	2024		47263	Cardiac resynchronization therapy implantable pacemaker	CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A071205	Pacemaker Found in Back-Up Mode		
									A0723	Unexpected Mode Switch		
							F01	Change in Therapeutic Response	A071205	Pacemaker Found in Back-Up Mode		
									A0723	Unexpected Mode Switch		
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A071205	Pacemaker Found in Back-Up Mode		
									A0723	Unexpected Mode Switch		
							F01	Change in Therapeutic Response	A071205	Pacemaker Found in Back-Up Mode		
									A0723	Unexpected Mode Switch		
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A071205	Pacemaker Found in Back-Up Mode		
									A0723	Unexpected Mode Switch		
							F01	Change in Therapeutic Response	A071205	Pacemaker Found in Back-Up Mode		
									A0723	Unexpected Mode Switch		
15045	2024	All other reportable incidents	35852	Single-chamber implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A072201	High Impedance		
							F26	No Health Consequences or Impact	A072201	High Impedance		
15046	2024	All other reportable incidents	37265	Dual-chamber implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E060101	Asystole	A070909	Oversensing		
									A0712	Pacing Problem		
									A090801	Signal Artifact/Noise		
							F1001	Absence of Treatment	A070909	Oversensing		
									A0712	Pacing Problem		
									A090801	Signal Artifact/Noise		
15047	2024	All other reportable incidents	47265	Dual-chamber implantable pacemaker, rate-responsive	CT1685	Cardiac pacemakers	E011903	Syncope/Fainting	A0203	Defective Device		
							F11	Non-Serious Injury/ Illness/ Impairment	A0203	Defective Device		
					CT257	Implantable cardiac pacemakers	E011903	Syncope/Fainting	A0203	Defective Device		
							F11	Non-Serious Injury/ Illness/ Impairment	A0203	Defective Device		
					CT2866	Implantable cardiac electrical stimulators	E011903	Syncope/Fainting	A0203	Defective Device		
							F11	Non-Serious Injury/ Illness/ Impairment	A0203	Defective Device		
15048	2024	All other reportable incidents	47265	Dual-chamber implantable pacemaker, rate-responsive	CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A070101	Failure to Capture		
							F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A070101	Failure to Capture		
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A070101	Failure to Capture		
							F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A070101	Failure to Capture		
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A070101	Failure to Capture		
							F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A070101	Failure to Capture		
15049	2024	All other reportable incidents	47270	Cardiac resynchronization therapy implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
15050	2024	All other reportable incidents	47265	Dual-chamber implantable pacemaker, rate-responsive	CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	13/11/2023	13/11/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	13/11/2023	13/11/2023
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	13/11/2023	13/11/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	13/11/2023	13/11/2023
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	13/11/2023	13/11/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	13/11/2023	13/11/2023
15051	2024	All other reportable incidents	35852	Single-chamber implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E011903	Syncope/Fainting	A0713	Defibrillation/Stimulation Problem		
							E060110	Ventricular Fibrillation	A0713	Defibrillation/Stimulation Problem		
							E2007	Fall	A0713	Defibrillation/Stimulation Problem		
							F08	Hospitalization or Prolonged Hospitalization	A0713	Defibrillation/Stimulation Problem		
							F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A0713	Defibrillation/Stimulation Problem		
							F1203	Life Threatening Illness or Injury	A0713	Defibrillation/Stimulation Problem		
							F2307	Reprogramming of Device	A0713	Defibrillation/Stimulation Problem		
15052	2024	All other reportable incidents	47265	Dual-chamber implantable pacemaker, rate-responsive	CT1685	Cardiac pacemakers	E0112	Dizziness	A0908	Incorrect, Inadequate or Imprecise Result or Readings	15/01/2024	15/01/2024
							F08	Hospitalization or Prolonged Hospitalization	A0908	Incorrect, Inadequate or Imprecise Result or Readings	15/01/2024	15/01/2024
							F1905	Device Revision or Replacement	A0908	Incorrect, Inadequate or Imprecise Result or Readings	15/01/2024	15/01/2024
					CT257	Implantable cardiac pacemakers	E0112	Dizziness	A0908	Incorrect, Inadequate or Imprecise Result or Readings	15/01/2024	15/01/2024
							F08	Hospitalization or Prolonged Hospitalization	A0908	Incorrect, Inadequate or Imprecise Result or Readings	15/01/2024	15/01/2024
							F1905	Device Revision or Replacement	A0908	Incorrect, Inadequate or Imprecise Result or Readings	15/01/2024	15/01/2024
					CT2866	Implantable cardiac electrical stimulators	E0112	Dizziness	A0908	Incorrect, Inadequate or Imprecise Result or Readings	15/01/2024	15/01/2024
							F08	Hospitalization or Prolonged Hospitalization	A0908	Incorrect, Inadequate or Imprecise Result or Readings	15/01/2024	15/01/2024
							F1905	Device Revision or Replacement	A0908	Incorrect, Inadequate or Imprecise Result or Readings	15/01/2024	15/01/2024
15053	2024	All other reportable incidents	47205	Cardiac pulse generator programmer	CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A13	Communication or Transmission Problem		
							F27	Problem Identified During Non-Clinical Procedure	A13	Communication or Transmission Problem		
15054	2024	All other reportable incidents	47265	Dual-chamber implantable pacemaker, rate-responsive	CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	10/09/2024	10/09/2024
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	10/09/2024	10/09/2024
							F12	Serious Injury/ Illness/ Impairment	A070504	Premature Discharge of Battery	10/09/2024	10/09/2024
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	10/09/2024	10/09/2024
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery	10/09/2024	10/09/2024
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	10/09/2024	10/09/2024
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	10/09/2024	10/09/2024
							F12	Serious Injury/ Illness/ Impairment	A070504	Premature Discharge of Battery	10/09/2024	10/09/2024
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	10/09/2024	10/09/2024
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery	10/09/2024	10/09/2024
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	10/09/2024	10/09/2024
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	10/09/2024	10/09/2024
							F12	Serious Injury/ Illness/ Impairment	A070504	Premature Discharge of Battery	10/09/2024	10/09/2024
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	10/09/2024	10/09/2024
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery	10/09/2024	10/09/2024
15055	2024	All other reportable incidents	47265	Dual-chamber implantable pacemaker, rate-responsive	CT1685	Cardiac pacemakers	E2311	Discomfort	A24	No Apparent Device Problem	23/11/2022	23/11/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	23/11/2022	23/11/2022
					CT257	Implantable cardiac pacemakers	E2311	Discomfort	A24	No Apparent Device Problem	23/11/2022	23/11/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	23/11/2022	23/11/2022

					CT2866	Implantable cardiac electrical stimulators	E2311	Discomfort	A24	No Apparent Device Problem	23/11/2022	23/11/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	23/11/2022	23/11/2022
15056	2024	All other reportable incidents	37265	Dual-chamber implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	21/03/2024	21/03/2024
							F1901	Additional Surgery	A070504	Premature Discharge of Battery	21/03/2024	21/03/2024
15057	2024	All other reportable incidents	47263	Cardiac resynchronization therapy implantable pacemaker	CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A0908	Incorrect, Inadequate or Imprecise Result or Readings	21/06/2023	21/06/2023
							F08	Hospitalization or Prolonged Hospitalization	A0908	Incorrect, Inadequate or Imprecise Result or Readings	21/06/2023	21/06/2023
							F1905	Device Revision or Replacement	A0908	Incorrect, Inadequate or Imprecise Result or Readings	21/06/2023	21/06/2023
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A0908	Incorrect, Inadequate or Imprecise Result or Readings	21/06/2023	21/06/2023
							F08	Hospitalization or Prolonged Hospitalization	A0908	Incorrect, Inadequate or Imprecise Result or Readings	21/06/2023	21/06/2023
							F1905	Device Revision or Replacement	A0908	Incorrect, Inadequate or Imprecise Result or Readings	21/06/2023	21/06/2023
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A0908	Incorrect, Inadequate or Imprecise Result or Readings	21/06/2023	21/06/2023
							F08	Hospitalization or Prolonged Hospitalization	A0908	Incorrect, Inadequate or Imprecise Result or Readings	21/06/2023	21/06/2023
							F1905	Device Revision or Replacement	A0908	Incorrect, Inadequate or Imprecise Result or Readings	21/06/2023	21/06/2023
15058	2024	All other reportable incidents	47263	Cardiac resynchronization therapy implantable pacemaker	CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A0908	Incorrect, Inadequate or Imprecise Result or Readings	15/06/2023	15/06/2023
							F08	Hospitalization or Prolonged Hospitalization	A0908	Incorrect, Inadequate or Imprecise Result or Readings	15/06/2023	15/06/2023
							F1905	Device Revision or Replacement	A0908	Incorrect, Inadequate or Imprecise Result or Readings	15/06/2023	15/06/2023
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A0908	Incorrect, Inadequate or Imprecise Result or Readings	15/06/2023	15/06/2023
							F08	Hospitalization or Prolonged Hospitalization	A0908	Incorrect, Inadequate or Imprecise Result or Readings	15/06/2023	15/06/2023
							F1905	Device Revision or Replacement	A0908	Incorrect, Inadequate or Imprecise Result or Readings	15/06/2023	15/06/2023
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A0908	Incorrect, Inadequate or Imprecise Result or Readings	15/06/2023	15/06/2023
							F08	Hospitalization or Prolonged Hospitalization	A0908	Incorrect, Inadequate or Imprecise Result or Readings	15/06/2023	15/06/2023
							F1905	Device Revision or Replacement	A0908	Incorrect, Inadequate or Imprecise Result or Readings	15/06/2023	15/06/2023
15059	2024	All other reportable incidents	47265	Dual-chamber implantable pacemaker, rate-responsive	CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A0908	Incorrect, Inadequate or Imprecise Result or Readings	27/09/2023	27/09/2023
							F08	Hospitalization or Prolonged Hospitalization	A0908	Incorrect, Inadequate or Imprecise Result or Readings	27/09/2023	27/09/2023
							F1905	Device Revision or Replacement	A0908	Incorrect, Inadequate or Imprecise Result or Readings	27/09/2023	27/09/2023
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A0908	Incorrect, Inadequate or Imprecise Result or Readings	27/09/2023	27/09/2023
							F08	Hospitalization or Prolonged Hospitalization	A0908	Incorrect, Inadequate or Imprecise Result or Readings	27/09/2023	27/09/2023
							F1905	Device Revision or Replacement	A0908	Incorrect, Inadequate or Imprecise Result or Readings	27/09/2023	27/09/2023
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A0908	Incorrect, Inadequate or Imprecise Result or Readings	27/09/2023	27/09/2023
							F08	Hospitalization or Prolonged Hospitalization	A0908	Incorrect, Inadequate or Imprecise Result or Readings	27/09/2023	27/09/2023
							F1905	Device Revision or Replacement	A0908	Incorrect, Inadequate or Imprecise Result or Readings	27/09/2023	27/09/2023
15060	2024	All other reportable incidents	47265	Dual-chamber implantable pacemaker, rate-responsive	CT1685	Cardiac pacemakers	E011903	Syncope/Fainting	A070504	Premature Discharge of Battery	29/01/2024	29/01/2024
									A0908	Incorrect, Inadequate or Imprecise Result or Readings	29/01/2024	29/01/2024
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	29/01/2024	29/01/2024
									A0908	Incorrect, Inadequate or Imprecise Result or Readings	29/01/2024	29/01/2024
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	29/01/2024	29/01/2024
									A0908	Incorrect, Inadequate or Imprecise Result or Readings	29/01/2024	29/01/2024
					CT257	Implantable cardiac pacemakers	E011903	Syncope/Fainting	A070504	Premature Discharge of Battery	29/01/2024	29/01/2024
									A0908	Incorrect, Inadequate or Imprecise Result or Readings	29/01/2024	29/01/2024
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	29/01/2024	29/01/2024
									A0908	Incorrect, Inadequate or Imprecise Result or Readings	29/01/2024	29/01/2024
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	29/01/2024	29/01/2024
									A0908	Incorrect, Inadequate or Imprecise Result or Readings	29/01/2024	29/01/2024
					CT2866	Implantable cardiac electrical stimulators	E011903	Syncope/Fainting	A070504	Premature Discharge of Battery	29/01/2024	29/01/2024
									A0908	Incorrect, Inadequate or Imprecise Result or Readings	29/01/2024	29/01/2024
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	29/01/2024	29/01/2024
									A0908	Incorrect, Inadequate or Imprecise Result or Readings	29/01/2024	29/01/2024
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	29/01/2024	29/01/2024
									A0908	Incorrect, Inadequate or Imprecise Result or Readings	29/01/2024	29/01/2024
15061	2024	All other reportable incidents			CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A072201	High Impedance		
							F26	No Health Consequences or Impact	A072201	High Impedance		
15062	2024	All other reportable incidents	47263	Cardiac resynchronization therapy implantable pacemaker	CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A0908	Incorrect, Inadequate or Imprecise Result or Readings	22/05/2023	22/05/2023
							F08	Hospitalization or Prolonged Hospitalization	A0908	Incorrect, Inadequate or Imprecise Result or Readings	22/05/2023	22/05/2023
							F1905	Device Revision or Replacement	A0908	Incorrect, Inadequate or Imprecise Result or Readings	22/05/2023	22/05/2023
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A0908	Incorrect, Inadequate or Imprecise Result or Readings	22/05/2023	22/05/2023
							F08	Hospitalization or Prolonged Hospitalization	A0908	Incorrect, Inadequate or Imprecise Result or Readings	22/05/2023	22/05/2023

						F1905	Device Revision or Replacement	A0908	Incorrect, Inadequate or Imprecise Result or Readings	22/05/2023	22/05/2023
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A0908	Incorrect, Inadequate or Imprecise Result or Readings	22/05/2023	22/05/2023
							F08	Hospitalization or Prolonged Hospitalization	A0908	Incorrect, Inadequate or Imprecise Result or Readings	22/05/2023	22/05/2023
							F1905	Device Revision or Replacement	A0908	Incorrect, Inadequate or Imprecise Result or Readings	22/05/2023	22/05/2023
15063	2024	All other reportable incidents	47265	Dual-chamber implantable pacemaker, rate-responsive	CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A0908	Incorrect, Inadequate or Imprecise Result or Readings	06/09/2023	06/09/2023
							F08	Hospitalization or Prolonged Hospitalization	A0908	Incorrect, Inadequate or Imprecise Result or Readings	06/09/2023	06/09/2023
							F1905	Device Revision or Replacement	A0908	Incorrect, Inadequate or Imprecise Result or Readings	06/09/2023	06/09/2023
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A0908	Incorrect, Inadequate or Imprecise Result or Readings	06/09/2023	06/09/2023
							F08	Hospitalization or Prolonged Hospitalization	A0908	Incorrect, Inadequate or Imprecise Result or Readings	06/09/2023	06/09/2023
							F1905	Device Revision or Replacement	A0908	Incorrect, Inadequate or Imprecise Result or Readings	06/09/2023	06/09/2023
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A0908	Incorrect, Inadequate or Imprecise Result or Readings	06/09/2023	06/09/2023
							F08	Hospitalization or Prolonged Hospitalization	A0908	Incorrect, Inadequate or Imprecise Result or Readings	06/09/2023	06/09/2023
							F1905	Device Revision or Replacement	A0908	Incorrect, Inadequate or Imprecise Result or Readings	06/09/2023	06/09/2023
15064	2024	All other reportable incidents	47267	Single-chamber implantable pacemaker, rate-responsive	CT1685	Cardiac pacemakers	E0112	Dizziness	A010402	Migration		
									A0712	Pacing Problem		
									A071202	Inaccurate Synchronization		
							E060109	Tachycardia	A010402	Migration		
									A0712	Pacing Problem		
									A071202	Inaccurate Synchronization		
							E0717	Dyspnea	A010402	Migration		
									A0712	Pacing Problem		
									A071202	Inaccurate Synchronization		
							F08	Hospitalization or Prolonged Hospitalization	A010402	Migration		
									A0712	Pacing Problem		
									A071202	Inaccurate Synchronization		
							F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A010402	Migration		
									A0712	Pacing Problem		
									A071202	Inaccurate Synchronization		
							F15	Recognized Device or Procedural Complication	A010402	Migration		
									A0712	Pacing Problem		
									A071202	Inaccurate Synchronization		
							F19	Surgical Intervention	A010402	Migration		
									A0712	Pacing Problem		
									A071202	Inaccurate Synchronization		
					CT257	Implantable cardiac pacemakers	E0112	Dizziness	A010402	Migration		
									A0712	Pacing Problem		
									A071202	Inaccurate Synchronization		
							E060109	Tachycardia	A010402	Migration		
									A0712	Pacing Problem		
									A071202	Inaccurate Synchronization		
							E0717	Dyspnea	A010402	Migration		
									A0712	Pacing Problem		
									A071202	Inaccurate Synchronization		
							F08	Hospitalization or Prolonged Hospitalization	A010402	Migration		
									A0712	Pacing Problem		
									A071202	Inaccurate Synchronization		
							F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A010402	Migration		
									A0712	Pacing Problem		
									A071202	Inaccurate Synchronization		
							F15	Recognized Device or Procedural Complication	A010402	Migration		
									A0712	Pacing Problem		
									A071202	Inaccurate Synchronization		
							F19	Surgical Intervention	A010402	Migration		
									A0712	Pacing Problem		
									A071202	Inaccurate Synchronization		
					CT2866	Implantable cardiac electrical stimulators	E0112	Dizziness	A010402	Migration		
									A0712	Pacing Problem		
									A071202	Inaccurate Synchronization		
							E060109	Tachycardia	A010402	Migration		
									A0712	Pacing Problem		
									A071202	Inaccurate Synchronization		
							E0717	Dyspnea	A010402	Migration		
									A0712	Pacing Problem		
									A071202	Inaccurate Synchronization		
							F08	Hospitalization or Prolonged Hospitalization	A010402	Migration		
									A0712	Pacing Problem		
									A071202	Inaccurate Synchronization		
							F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A010402	Migration		
									A0712	Pacing Problem		
									A071202	Inaccurate Synchronization		
							F15	Recognized Device or Procedural Complication	A010402	Migration		
									A0712	Pacing Problem		
									A071202	Inaccurate Synchronization		
							F19	Surgical Intervention	A010402	Migration		
									A0712	Pacing Problem		
									A071202	Inaccurate Synchronization		
15065	2024	All other reportable incidents	47270	Cardiac resynchronization therapy implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		

15066	2024	All other reportable incidents	47265	Dual-chamber implantable pacemaker, rate-responsive	CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A0908	Incorrect, Inadequate or Imprecise Result or Readings	04/08/2023	04/08/2023
							F08	Hospitalization or Prolonged Hospitalization	A0908	Incorrect, Inadequate or Imprecise Result or Readings	04/08/2023	04/08/2023
							F1905	Device Revision or Replacement	A0908	Incorrect, Inadequate or Imprecise Result or Readings	04/08/2023	04/08/2023
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A0908	Incorrect, Inadequate or Imprecise Result or Readings	04/08/2023	04/08/2023
							F08	Hospitalization or Prolonged Hospitalization	A0908	Incorrect, Inadequate or Imprecise Result or Readings	04/08/2023	04/08/2023
							F1905	Device Revision or Replacement	A0908	Incorrect, Inadequate or Imprecise Result or Readings	04/08/2023	04/08/2023
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A0908	Incorrect, Inadequate or Imprecise Result or Readings	04/08/2023	04/08/2023
							F08	Hospitalization or Prolonged Hospitalization	A0908	Incorrect, Inadequate or Imprecise Result or Readings	04/08/2023	04/08/2023
							F1905	Device Revision or Replacement	A0908	Incorrect, Inadequate or Imprecise Result or Readings	04/08/2023	04/08/2023
15067	2024		47265	Dual-chamber implantable pacemaker, rate-responsive	CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	21/11/2023	21/11/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	21/11/2023	21/11/2023
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	21/11/2023	21/11/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	21/11/2023	21/11/2023
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	21/11/2023	21/11/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	21/11/2023	21/11/2023
15068	2024		37265	Dual-chamber implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E0601	Arrhythmia	A07	Electrical /Electronic Property Problem		
									A071201	Failure to Convert Rhythm		
									A071302	Inappropriate/Inadequate Shock/Stimulation		
							F02	Death	A07	Electrical /Electronic Property Problem		
									A071201	Failure to Convert Rhythm		
									A071302	Inappropriate/Inadequate Shock/Stimulation		
15069	2024	All other reportable incidents	17882	Physiologic-monitoring defibrillation system	CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070908	Failure to Sense		
							F26	No Health Consequences or Impact	A070908	Failure to Sense		
15070	2024	All other reportable incidents	47266	Single-chamber implantable pacemaker, fixed-rate/demand	CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A070909	Oversensing		
							F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A070909	Oversensing		
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A070909	Oversensing		
							F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A070909	Oversensing		
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A070909	Oversensing		
							F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A070909	Oversensing		
15071	2024	All other reportable incidents	47265	Dual-chamber implantable pacemaker, rate-responsive	CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem	01/12/2023	01/12/2023
									A070504	Premature Discharge of Battery	01/12/2023	01/12/2023
							F08	Hospitalization or Prolonged Hospitalization	A0705	Battery Problem	01/12/2023	01/12/2023
									A070504	Premature Discharge of Battery	01/12/2023	01/12/2023
							F1903	Device Explantation	A0705	Battery Problem	01/12/2023	01/12/2023
									A070504	Premature Discharge of Battery	01/12/2023	01/12/2023
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem	01/12/2023	01/12/2023
									A070504	Premature Discharge of Battery	01/12/2023	01/12/2023
							F08	Hospitalization or Prolonged Hospitalization	A0705	Battery Problem	01/12/2023	01/12/2023
									A070504	Premature Discharge of Battery	01/12/2023	01/12/2023
							F1903	Device Explantation	A0705	Battery Problem	01/12/2023	01/12/2023
									A070504	Premature Discharge of Battery	01/12/2023	01/12/2023
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem	01/12/2023	01/12/2023
									A070504	Premature Discharge of Battery	01/12/2023	01/12/2023
							F08	Hospitalization or Prolonged Hospitalization	A0705	Battery Problem	01/12/2023	01/12/2023
									A070504	Premature Discharge of Battery	01/12/2023	01/12/2023
							F1903	Device Explantation	A0705	Battery Problem	01/12/2023	01/12/2023
									A070504	Premature Discharge of Battery	01/12/2023	01/12/2023
15072	2024	All other reportable incidents	47265	Dual-chamber implantable pacemaker, rate-responsive	CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	21/11/2023	21/11/2023
									A160601	Premature Elective Replacement Indicator	21/11/2023	21/11/2023
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	21/11/2023	21/11/2023
									A160601	Premature Elective Replacement Indicator	21/11/2023	21/11/2023
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	21/11/2023	21/11/2023
									A160601	Premature Elective Replacement Indicator	21/11/2023	21/11/2023
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	21/11/2023	21/11/2023
									A160601	Premature Elective Replacement Indicator	21/11/2023	21/11/2023
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	21/11/2023	21/11/2023
									A160601	Premature Elective Replacement Indicator	21/11/2023	21/11/2023
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	21/11/2023	21/11/2023
									A160601	Premature Elective Replacement Indicator	21/11/2023	21/11/2023
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	21/11/2023	21/11/2023
									A160601	Premature Elective Replacement Indicator	21/11/2023	21/11/2023
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	21/11/2023	21/11/2023
									A160601	Premature Elective Replacement Indicator	21/11/2023	21/11/2023
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	21/11/2023	21/11/2023
									A160601	Premature Elective Replacement Indicator	21/11/2023	21/11/2023
15073	2024	All other reportable incidents	47265	Dual-chamber implantable pacemaker, rate-responsive	CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem		
							F26	No Health Consequences or Impact	A0705	Battery Problem		
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem		
							F26	No Health Consequences or Impact	A0705	Battery Problem		
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem		
							F26	No Health Consequences or Impact	A0705	Battery Problem		
15074	2024	All other reportable incidents	47265	Dual-chamber implantable pacemaker, rate-responsive	CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	31/01/2024	31/01/2024
							F1901	Additional Surgery	A24	No Apparent Device Problem	31/01/2024	31/01/2024
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	31/01/2024	31/01/2024
							F1901	Additional Surgery	A24	No Apparent Device Problem	31/01/2024	31/01/2024
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	31/01/2024	31/01/2024
							F1901	Additional Surgery	A24	No Apparent Device Problem	31/01/2024	31/01/2024
15075	2024		47265	Dual-chamber implantable pacemaker, rate-responsive	CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	11/12/2023	11/12/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	11/12/2023	11/12/2023
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	11/12/2023	11/12/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	11/12/2023	11/12/2023

					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	11/12/2023	11/12/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	11/12/2023	11/12/2023
15076	2024	All other reportable incidents	17882	Physiologic-monitoring defibrillation system	CT1684	Cardiac pacemakers and associated devices	F26	No Health Consequences or Impact	A071302	Inappropriate/Inadequate Shock/Stimulation		
15077	2024	All other reportable incidents	47265	Dual-chamber implantable pacemaker, rate-responsive	CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
15078	2024	All other reportable incidents	47265	Dual-chamber implantable pacemaker, rate-responsive	CT1685	Cardiac pacemakers	E0112	Dizziness	A070909	Oversensing	27/03/2024	27/03/2024
									A0712	Pacing Problem	27/03/2024	27/03/2024
									A0908	Incorrect, Inadequate or Imprecise Result or Readings	27/03/2024	27/03/2024
									A090801	Signal Artifact/Noise	27/03/2024	27/03/2024
							E020201	Anxiety	A070909	Oversensing	27/03/2024	27/03/2024
									A0712	Pacing Problem	27/03/2024	27/03/2024
									A0908	Incorrect, Inadequate or Imprecise Result or Readings	27/03/2024	27/03/2024
									A090801	Signal Artifact/Noise	27/03/2024	27/03/2024
							E0717	Dyspnea	A070909	Oversensing	27/03/2024	27/03/2024
									A0712	Pacing Problem	27/03/2024	27/03/2024
									A0908	Incorrect, Inadequate or Imprecise Result or Readings	27/03/2024	27/03/2024
									A090801	Signal Artifact/Noise	27/03/2024	27/03/2024
							F08	Hospitalization or Prolonged Hospitalization	A070909	Oversensing	27/03/2024	27/03/2024
									A0712	Pacing Problem	27/03/2024	27/03/2024
									A0908	Incorrect, Inadequate or Imprecise Result or Readings	27/03/2024	27/03/2024
									A090801	Signal Artifact/Noise	27/03/2024	27/03/2024
							F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A070909	Oversensing	27/03/2024	27/03/2024
									A0712	Pacing Problem	27/03/2024	27/03/2024
									A0908	Incorrect, Inadequate or Imprecise Result or Readings	27/03/2024	27/03/2024
									A090801	Signal Artifact/Noise	27/03/2024	27/03/2024
							F1905	Device Revision or Replacement	A070909	Oversensing	27/03/2024	27/03/2024
									A0712	Pacing Problem	27/03/2024	27/03/2024
									A0908	Incorrect, Inadequate or Imprecise Result or Readings	27/03/2024	27/03/2024
									A090801	Signal Artifact/Noise	27/03/2024	27/03/2024
					CT257	Implantable cardiac pacemakers	E0112	Dizziness	A070909	Oversensing	27/03/2024	27/03/2024
									A0712	Pacing Problem	27/03/2024	27/03/2024
									A0908	Incorrect, Inadequate or Imprecise Result or Readings	27/03/2024	27/03/2024
									A090801	Signal Artifact/Noise	27/03/2024	27/03/2024
							E020201	Anxiety	A070909	Oversensing	27/03/2024	27/03/2024
									A0712	Pacing Problem	27/03/2024	27/03/2024
									A0908	Incorrect, Inadequate or Imprecise Result or Readings	27/03/2024	27/03/2024
									A090801	Signal Artifact/Noise	27/03/2024	27/03/2024
							E0717	Dyspnea	A070909	Oversensing	27/03/2024	27/03/2024
									A0712	Pacing Problem	27/03/2024	27/03/2024
									A0908	Incorrect, Inadequate or Imprecise Result or Readings	27/03/2024	27/03/2024
									A090801	Signal Artifact/Noise	27/03/2024	27/03/2024
							F08	Hospitalization or Prolonged Hospitalization	A070909	Oversensing	27/03/2024	27/03/2024
									A0712	Pacing Problem	27/03/2024	27/03/2024
									A0908	Incorrect, Inadequate or Imprecise Result or Readings	27/03/2024	27/03/2024
									A090801	Signal Artifact/Noise	27/03/2024	27/03/2024
							F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A070909	Oversensing	27/03/2024	27/03/2024
									A0712	Pacing Problem	27/03/2024	27/03/2024
									A0908	Incorrect, Inadequate or Imprecise Result or Readings	27/03/2024	27/03/2024
									A090801	Signal Artifact/Noise	27/03/2024	27/03/2024
							F1905	Device Revision or Replacement	A070909	Oversensing	27/03/2024	27/03/2024
									A0712	Pacing Problem	27/03/2024	27/03/2024
									A0908	Incorrect, Inadequate or Imprecise Result or Readings	27/03/2024	27/03/2024
									A090801	Signal Artifact/Noise	27/03/2024	27/03/2024
					CT2866	Implantable cardiac electrical stimulators	E0112	Dizziness	A070909	Oversensing	27/03/2024	27/03/2024
									A0712	Pacing Problem	27/03/2024	27/03/2024
									A0908	Incorrect, Inadequate or Imprecise Result or Readings	27/03/2024	27/03/2024
									A090801	Signal Artifact/Noise	27/03/2024	27/03/2024
							E020201	Anxiety	A070909	Oversensing	27/03/2024	27/03/2024
									A0712	Pacing Problem	27/03/2024	27/03/2024
									A0908	Incorrect, Inadequate or Imprecise Result or Readings	27/03/2024	27/03/2024
									A090801	Signal Artifact/Noise	27/03/2024	27/03/2024
							E0717	Dyspnea	A070909	Oversensing	27/03/2024	27/03/2024
									A0712	Pacing Problem	27/03/2024	27/03/2024
									A0908	Incorrect, Inadequate or Imprecise Result or Readings	27/03/2024	27/03/2024
									A090801	Signal Artifact/Noise	27/03/2024	27/03/2024
							F08	Hospitalization or Prolonged Hospitalization	A070909	Oversensing	27/03/2024	27/03/2024
									A0712	Pacing Problem	27/03/2024	27/03/2024
									A0908	Incorrect, Inadequate or Imprecise Result or Readings	27/03/2024	27/03/2024
									A090801	Signal Artifact/Noise	27/03/2024	27/03/2024
							F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A070909	Oversensing	27/03/2024	27/03/2024
									A0712	Pacing Problem	27/03/2024	27/03/2024

									A0908	Incorrect, Inadequate or Imprecise Result or Readings	27/03/2024	27/03/2024
									A090801	Signal Artifact/Noise	27/03/2024	27/03/2024
							F1905	Device Revision or Replacement	A070909	Oversensing	27/03/2024	27/03/2024
									A0712	Pacing Problem	27/03/2024	27/03/2024
									A0908	Incorrect, Inadequate or Imprecise Result or Readings	27/03/2024	27/03/2024
									A090801	Signal Artifact/Noise	27/03/2024	27/03/2024
15079	2024	All other reportable incidents	47270	Cardiac resynchronization therapy implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E210401	Shock from Patient Lead(s)	A070909	Oversensing		
									A070910	Undersensing		
							F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A070909	Oversensing		
									A070910	Undersensing		
							F17	Sedation	A070909	Oversensing		
									A070910	Undersensing		
							F2203	Imaging Required	A070909	Oversensing		
									A070910	Undersensing		
15080	2024	All other reportable incidents	47205	Cardiac pulse generator programmer	CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A13	Communication or Transmission Problem		
							F26	No Health Consequences or Impact	A13	Communication or Transmission Problem		
15081	2024	All other reportable incidents	35852	Single-chamber implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A0405	Degraded		
							F27	Problem Identified During Non-Clinical Procedure	A0405	Degraded		
15082	2024	All other reportable incidents	17882	Physiologic-monitoring defibrillation system	CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F27	Problem Identified During Non-Clinical Procedure	A070504	Premature Discharge of Battery		
15083	2024	All other reportable incidents	35852	Single-chamber implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E060102	Atrial Fibrillation	A27	Appropriate Device Problem Term/Code not Available		
							F12	Serious Injury/ Illness/ Impairment	A27	Appropriate Device Problem Term/Code not Available		
							F23	Unexpected Medical Intervention	A27	Appropriate Device Problem Term/Code not Available		
							F2303	Medication Required	A27	Appropriate Device Problem Term/Code not Available		
15084	2024	All other reportable incidents	47205	Cardiac pulse generator programmer	CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A13	Communication or Transmission Problem		
							F26	No Health Consequences or Impact	A13	Communication or Transmission Problem		
15085	2024	All other reportable incidents	47263	Cardiac resynchronization therapy implantable pacemaker	CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem	25/03/2024	25/03/2024
									A070504	Premature Discharge of Battery	25/03/2024	25/03/2024
							F08	Hospitalization or Prolonged Hospitalization	A0705	Battery Problem	25/03/2024	25/03/2024
									A070504	Premature Discharge of Battery	25/03/2024	25/03/2024
							F1905	Device Revision or Replacement	A0705	Battery Problem	25/03/2024	25/03/2024
									A070504	Premature Discharge of Battery	25/03/2024	25/03/2024
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem	25/03/2024	25/03/2024
									A070504	Premature Discharge of Battery	25/03/2024	25/03/2024
							F08	Hospitalization or Prolonged Hospitalization	A0705	Battery Problem	25/03/2024	25/03/2024
									A070504	Premature Discharge of Battery	25/03/2024	25/03/2024
							F1905	Device Revision or Replacement	A0705	Battery Problem	25/03/2024	25/03/2024
									A070504	Premature Discharge of Battery	25/03/2024	25/03/2024
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem	25/03/2024	25/03/2024
									A070504	Premature Discharge of Battery	25/03/2024	25/03/2024
							F08	Hospitalization or Prolonged Hospitalization	A0705	Battery Problem	25/03/2024	25/03/2024
									A070504	Premature Discharge of Battery	25/03/2024	25/03/2024
							F1905	Device Revision or Replacement	A0705	Battery Problem	25/03/2024	25/03/2024
									A070504	Premature Discharge of Battery	25/03/2024	25/03/2024
15086	2024	All other reportable incidents	35852	Single-chamber implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E210401	Shock from Patient Lead(s)	A0705	Battery Problem		
									A070602	Delayed Charge Time		
									A071302	Inappropriate/Inadequate Shock/Stimulation		
							F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A0705	Battery Problem		
									A070602	Delayed Charge Time		
									A071302	Inappropriate/Inadequate Shock/Stimulation		
15087	2024		47265	Dual-chamber implantable pacemaker, rate-responsive	CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem		
							F1905	Device Revision or Replacement	A0705	Battery Problem		
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem		
							F1905	Device Revision or Replacement	A0705	Battery Problem		
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem		
							F1905	Device Revision or Replacement	A0705	Battery Problem		
15088	2024	All other reportable incidents	47267	Single-chamber implantable pacemaker, rate-responsive	CT1685	Cardiac pacemakers	E2103	Device Overstimulation of Tissue	A0908	Incorrect, Inadequate or Imprecise Result or Readings	28/09/2023	28/09/2023
							F08	Hospitalization or Prolonged Hospitalization	A0908	Incorrect, Inadequate or Imprecise Result or Readings	28/09/2023	28/09/2023
							F1905	Device Revision or Replacement	A0908	Incorrect, Inadequate or Imprecise Result or Readings	28/09/2023	28/09/2023
					CT257	Implantable cardiac pacemakers	E2103	Device Overstimulation of Tissue	A0908	Incorrect, Inadequate or Imprecise Result or Readings	28/09/2023	28/09/2023
							F08	Hospitalization or Prolonged Hospitalization	A0908	Incorrect, Inadequate or Imprecise Result or Readings	28/09/2023	28/09/2023
							F1905	Device Revision or Replacement	A0908	Incorrect, Inadequate or Imprecise Result or Readings	28/09/2023	28/09/2023
					CT2866	Implantable cardiac electrical stimulators	E2103	Device Overstimulation of Tissue	A0908	Incorrect, Inadequate or Imprecise Result or Readings	28/09/2023	28/09/2023
							F08	Hospitalization or Prolonged Hospitalization	A0908	Incorrect, Inadequate or Imprecise Result or Readings	28/09/2023	28/09/2023
							F1905	Device Revision or Replacement	A0908	Incorrect, Inadequate or Imprecise Result or Readings	28/09/2023	28/09/2023
15089	2024		47265	Dual-chamber implantable pacemaker, rate-responsive	CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	28/03/2024	28/03/2024
							F1901	Additional Surgery	A24	No Apparent Device Problem	28/03/2024	28/03/2024
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	28/03/2024	28/03/2024
							F1901	Additional Surgery	A24	No Apparent Device Problem	28/03/2024	28/03/2024
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	28/03/2024	28/03/2024
							F1901	Additional Surgery	A24	No Apparent Device Problem	28/03/2024	28/03/2024
15090	2024	All other reportable incidents	17882	Physiologic-monitoring defibrillation system	CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A1603	Failure of Device to Self-Test		
							F27	Problem Identified During Non-Clinical Procedure	A1603	Failure of Device to Self-Test		
15091	2024	All other reportable incidents	17882	Physiologic-monitoring defibrillation system	CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A1601	Device Alarm System		
							F27	Problem Identified During Non-Clinical Procedure	A1601	Device Alarm System		
15092	2024	All other reportable incidents	35224	External pacemaker, epicardial pacing	CT1684	Cardiac pacemakers and associated devices	E060110	Ventricular Fibrillation	A0709	Device Sensing Problem		

									A09	Output Problem			
							F1203	Life Threatening Illness or Injury	A0709	Device Sensing Problem			
									A09	Output Problem			
							F23	Unexpected Medical Intervention	A0709	Device Sensing Problem			
									A09	Output Problem			
					CT1685	Cardiac pacemakers	E060110	Ventricular Fibrillation	A0709	Device Sensing Problem			
									A09	Output Problem			
							F1203	Life Threatening Illness or Injury	A0709	Device Sensing Problem			
									A09	Output Problem			
							F23	Unexpected Medical Intervention	A0709	Device Sensing Problem			
									A09	Output Problem			
15093	2024	All other reportable incidents	47265	Dual-chamber implantable pacemaker, rate-responsive	CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem			
									A070504	Premature Discharge of Battery			
							F26	No Health Consequences or Impact	A0705	Battery Problem			
									A070504	Premature Discharge of Battery			
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem			
									A070504	Premature Discharge of Battery			
							F26	No Health Consequences or Impact	A0705	Battery Problem			
									A070504	Premature Discharge of Battery			
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem			
									A070504	Premature Discharge of Battery			
							F26	No Health Consequences or Impact	A0705	Battery Problem			
									A070504	Premature Discharge of Battery			
15094	2024	All other reportable incidents	47265	Dual-chamber implantable pacemaker, rate-responsive	CT1685	Cardiac pacemakers	E0717	Dyspnea	A0712	Pacing Problem			
							E1621	Muscle Weakness/Atrophy	A0712	Pacing Problem			
							F2307	Reprogramming of Device	A0712	Pacing Problem			
					CT257	Implantable cardiac pacemakers	E0717	Dyspnea	A0712	Pacing Problem			
							E1621	Muscle Weakness/Atrophy	A0712	Pacing Problem			
							F2307	Reprogramming of Device	A0712	Pacing Problem			
					CT2866	Implantable cardiac electrical stimulators	E0717	Dyspnea	A0712	Pacing Problem			
							E1621	Muscle Weakness/Atrophy	A0712	Pacing Problem			
							F2307	Reprogramming of Device	A0712	Pacing Problem			
15095	2024	All other reportable incidents	35852	Single-chamber implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A0405	Degraded			
							F27	Problem Identified During Non-Clinical Procedure	A0405	Degraded			
15096	2024	All other reportable incidents	17882	Physiologic-monitoring defibrillation system	CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A0719	Unexpected Shutdown			
							F27	Problem Identified During Non-Clinical Procedure	A0719	Unexpected Shutdown			
15097	2024	All other reportable incidents	17882	Physiologic-monitoring defibrillation system	CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem			
							F27	Problem Identified During Non-Clinical Procedure	A0705	Battery Problem			
15098	2024	All other reportable incidents	47265	Dual-chamber implantable pacemaker, rate-responsive	CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery			
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery			
							F12	Serious Injury/ Illness/ Impairment	A070504	Premature Discharge of Battery			
							F1903	Device Explantation	A070504	Premature Discharge of Battery			
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery			
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery			
							F12	Serious Injury/ Illness/ Impairment	A070504	Premature Discharge of Battery			
							F1903	Device Explantation	A070504	Premature Discharge of Battery			
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery			
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery			
							F12	Serious Injury/ Illness/ Impairment	A070504	Premature Discharge of Battery			
							F1903	Device Explantation	A070504	Premature Discharge of Battery			
15099	2024		47265	Dual-chamber implantable pacemaker, rate-responsive	CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	03/04/2024	03/04/2024	
							F1901	Additional Surgery	A24	No Apparent Device Problem	03/04/2024	03/04/2024	
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	03/04/2024	03/04/2024	
							F1901	Additional Surgery	A24	No Apparent Device Problem	03/04/2024	03/04/2024	
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	03/04/2024	03/04/2024	
							F1901	Additional Surgery	A24	No Apparent Device Problem	03/04/2024	03/04/2024	
15100	2024	All other reportable incidents	47205	Cardiac pulse generator programmer	CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A13	Communication or Transmission Problem			
							F26	No Health Consequences or Impact	A13	Communication or Transmission Problem			
15101	2024	All other reportable incidents	47265	Dual-chamber implantable pacemaker, rate-responsive	CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A0908	Incorrect, Inadequate or Imprecise Result or Readings			
									A1304	Telemetry Discrepancy			
							F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A0908	Incorrect, Inadequate or Imprecise Result or Readings			
									A1304	Telemetry Discrepancy			
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A0908	Incorrect, Inadequate or Imprecise Result or Readings			
									A1304	Telemetry Discrepancy			
							F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A0908	Incorrect, Inadequate or Imprecise Result or Readings			
									A1304	Telemetry Discrepancy			
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A0908	Incorrect, Inadequate or Imprecise Result or Readings			
									A1304	Telemetry Discrepancy			
							F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A0908	Incorrect, Inadequate or Imprecise Result or Readings			
									A1304	Telemetry Discrepancy			
15102	2024	All other reportable incidents	35852	Single-chamber implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A0908	Incorrect, Inadequate or Imprecise Result or Readings			
							F27	Problem Identified During Non-Clinical Procedure	A0908	Incorrect, Inadequate or Imprecise Result or Readings			
15103	2024	All other reportable incidents	17882	Physiologic-monitoring defibrillation system	CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070803	Failure to Power Up			
							F27	Problem Identified During Non-Clinical Procedure	A070803	Failure to Power Up			
15104	2024	All other reportable incidents	35852	Single-chamber implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E060109	Tachycardia	A0508	Noise, Audible			
									A070909	Oversensing			
									A070910	Undersensing			
									A0713	Defibrillation/Stimulation Problem			
							E0623	Unspecified Heart Problem	A0508	Noise, Audible			
									A070909	Oversensing			
									A070910	Undersensing			
									A0713	Defibrillation/Stimulation Problem			

							F08	Hospitalization or Prolonged Hospitalization	A0508	Noise, Audible		
									A070909	Oversensing		
									A070910	Undersensing		
									A0713	Defibrillation/Stimulation Problem		
							F15	Recognized Device or Procedural Complication	A0508	Noise, Audible		
									A070909	Oversensing		
									A070910	Undersensing		
									A0713	Defibrillation/Stimulation Problem		
							F2303	Medication Required	A0508	Noise, Audible		
									A070909	Oversensing		
									A070910	Undersensing		
									A0713	Defibrillation/Stimulation Problem		
15105	2024	All other reportable incidents	35852	Single-chamber implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A0712	Pacing Problem		
							F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A0712	Pacing Problem		
15106	2024		17882	Physiologic-monitoring defibrillation system	CT1684	Cardiac pacemakers and associated devices	E2401	Insufficient Information	A150101	Activation Failure		
							F05	Delay to Treatment/ Therapy	A150101	Activation Failure		
15107	2024	All other reportable incidents	37265	Dual-chamber implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A070909	Oversensing		
									A071204	No Pacing		
							F22	Unexpected Diagnostic Intervention	A070909	Oversensing		
									A071204	No Pacing		
15108	2024	All other reportable incidents	36106	Implantable pacemaker analyser, intraoperative	CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A1104	Computer Operating System Problem		
							F26	No Health Consequences or Impact	A1104	Computer Operating System Problem		
15109	2024	All other reportable incidents	35852	Single-chamber implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem	27/09/2023	27/09/2023
									A070504	Premature Discharge of Battery	27/09/2023	27/09/2023
									A0908	Incorrect, Inadequate or Imprecise Result or Readings	27/09/2023	27/09/2023
							F08	Hospitalization or Prolonged Hospitalization	A0705	Battery Problem	27/09/2023	27/09/2023
									A070504	Premature Discharge of Battery	27/09/2023	27/09/2023
									A0908	Incorrect, Inadequate or Imprecise Result or Readings	27/09/2023	27/09/2023
							F1905	Device Revision or Replacement	A0705	Battery Problem	27/09/2023	27/09/2023
									A070504	Premature Discharge of Battery	27/09/2023	27/09/2023
									A0908	Incorrect, Inadequate or Imprecise Result or Readings	27/09/2023	27/09/2023
15110	2024		35224	External pacemaker, epicardial pacing	CT1685	Cardiac pacemakers	E060107	Idioventricular Rhythm	A01	Patient Device Interaction Problem		
									A0202	Defective Component		
							E060109	Tachycardia	A01	Patient Device Interaction Problem		
									A0202	Defective Component		
							F15	Recognized Device or Procedural Complication	A01	Patient Device Interaction Problem		
									A0202	Defective Component		
15111	2024	All other reportable incidents	35852	Single-chamber implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
									A0908	Incorrect, Inadequate or Imprecise Result or Readings		
							F24	Insufficient Health Impact Information	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
									A0908	Incorrect, Inadequate or Imprecise Result or Readings		
15112	2024	All other reportable incidents	12913	Pacemaker, implantable	CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A27	Appropriate Device Problem Term/Code not Available	01/01/2014	31/12/2014
							F08	Hospitalization or Prolonged Hospitalization	A27	Appropriate Device Problem Term/Code not Available	01/01/2014	31/12/2014
							F12	Serious Injury/ Illness/ Impairment	A27	Appropriate Device Problem Term/Code not Available	01/01/2014	31/12/2014
							F1903	Device Explantation	A27	Appropriate Device Problem Term/Code not Available	01/01/2014	31/12/2014
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A27	Appropriate Device Problem Term/Code not Available	01/01/2014	31/12/2014
							F08	Hospitalization or Prolonged Hospitalization	A27	Appropriate Device Problem Term/Code not Available	01/01/2014	31/12/2014
							F12	Serious Injury/ Illness/ Impairment	A27	Appropriate Device Problem Term/Code not Available	01/01/2014	31/12/2014
							F1903	Device Explantation	A27	Appropriate Device Problem Term/Code not Available	01/01/2014	31/12/2014
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A27	Appropriate Device Problem Term/Code not Available	01/01/2014	31/12/2014
							F08	Hospitalization or Prolonged Hospitalization	A27	Appropriate Device Problem Term/Code not Available	01/01/2014	31/12/2014
							F12	Serious Injury/ Illness/ Impairment	A27	Appropriate Device Problem Term/Code not Available	01/01/2014	31/12/2014
							F1903	Device Explantation	A27	Appropriate Device Problem Term/Code not Available	01/01/2014	31/12/2014
15113	2024	All other reportable incidents	47265	Dual-chamber implantable pacemaker, rate-responsive	CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A0908	Incorrect, Inadequate or Imprecise Result or Readings	18/12/2023	18/12/2023
							F08	Hospitalization or Prolonged Hospitalization	A0908	Incorrect, Inadequate or Imprecise Result or Readings	18/12/2023	18/12/2023
							F1905	Device Revision or Replacement	A0908	Incorrect, Inadequate or Imprecise Result or Readings	18/12/2023	18/12/2023
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A0908	Incorrect, Inadequate or Imprecise Result or Readings	18/12/2023	18/12/2023
							F08	Hospitalization or Prolonged Hospitalization	A0908	Incorrect, Inadequate or Imprecise Result or Readings	18/12/2023	18/12/2023
							F1905	Device Revision or Replacement	A0908	Incorrect, Inadequate or Imprecise Result or Readings	18/12/2023	18/12/2023
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A0908	Incorrect, Inadequate or Imprecise Result or Readings	18/12/2023	18/12/2023
							F08	Hospitalization or Prolonged Hospitalization	A0908	Incorrect, Inadequate or Imprecise Result or Readings	18/12/2023	18/12/2023
							F1905	Device Revision or Replacement	A0908	Incorrect, Inadequate or Imprecise Result or Readings	18/12/2023	18/12/2023
15114	2024	All other reportable incidents	37265	Dual-chamber implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E0112	Dizziness	A070909	Oversensing		
									A0712	Pacing Problem		
							E060104	Bradycardia	A070909	Oversensing		
									A0712	Pacing Problem		

							F10	Inadequate/inappropriate Treatment or Diagnostic Exposure	A070909	Oversensing		
									A0712	Pacing Problem		
							F2203	Imaging Required	A070909	Oversensing		
									A0712	Pacing Problem		
15115	2024	All other reportable incidents	47265	Dual-chamber implantable pacemaker, rate-responsive	CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A0908	Incorrect, Inadequate or Imprecise Result or Readings	15/04/2024	15/04/2024
							F08	Hospitalization or Prolonged Hospitalization	A0908	Incorrect, Inadequate or Imprecise Result or Readings	15/04/2024	15/04/2024
							F1905	Device Revision or Replacement	A0908	Incorrect, Inadequate or Imprecise Result or Readings	15/04/2024	15/04/2024
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A0908	Incorrect, Inadequate or Imprecise Result or Readings	15/04/2024	15/04/2024
							F08	Hospitalization or Prolonged Hospitalization	A0908	Incorrect, Inadequate or Imprecise Result or Readings	15/04/2024	15/04/2024
							F1905	Device Revision or Replacement	A0908	Incorrect, Inadequate or Imprecise Result or Readings	15/04/2024	15/04/2024
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A0908	Incorrect, Inadequate or Imprecise Result or Readings	15/04/2024	15/04/2024
							F08	Hospitalization or Prolonged Hospitalization	A0908	Incorrect, Inadequate or Imprecise Result or Readings	15/04/2024	15/04/2024
							F1905	Device Revision or Replacement	A0908	Incorrect, Inadequate or Imprecise Result or Readings	15/04/2024	15/04/2024
15116	2024	All other reportable incidents	47265	Dual-chamber implantable pacemaker, rate-responsive	CT1685	Cardiac pacemakers	E020203	Distress	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
							F24	Insufficient Health Impact Information	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
					CT257	Implantable cardiac pacemakers	E020203	Distress	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
							F24	Insufficient Health Impact Information	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
					CT2866	Implantable cardiac electrical stimulators	E020203	Distress	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
							F24	Insufficient Health Impact Information	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
15117	2024	All other reportable incidents	47263	Cardiac resynchronization therapy implantable pacemaker	CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A0908	Incorrect, Inadequate or Imprecise Result or Readings		
							F26	No Health Consequences or Impact	A0908	Incorrect, Inadequate or Imprecise Result or Readings		
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A0908	Incorrect, Inadequate or Imprecise Result or Readings		
							F26	No Health Consequences or Impact	A0908	Incorrect, Inadequate or Imprecise Result or Readings		
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A0908	Incorrect, Inadequate or Imprecise Result or Readings		
							F26	No Health Consequences or Impact	A0908	Incorrect, Inadequate or Imprecise Result or Readings		
15118	2024	All other reportable incidents	47265	Dual-chamber implantable pacemaker, rate-responsive	CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A0711	Interrogation Problem	12/04/2024	12/04/2024
							F08	Hospitalization or Prolonged Hospitalization	A0711	Interrogation Problem	12/04/2024	12/04/2024
							F1905	Device Revision or Replacement	A0711	Interrogation Problem	12/04/2024	12/04/2024
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A0711	Interrogation Problem	12/04/2024	12/04/2024
							F08	Hospitalization or Prolonged Hospitalization	A0711	Interrogation Problem	12/04/2024	12/04/2024
							F1905	Device Revision or Replacement	A0711	Interrogation Problem	12/04/2024	12/04/2024
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A0711	Interrogation Problem	12/04/2024	12/04/2024
							F08	Hospitalization or Prolonged Hospitalization	A0711	Interrogation Problem	12/04/2024	12/04/2024
							F1905	Device Revision or Replacement	A0711	Interrogation Problem	12/04/2024	12/04/2024
15119	2024	All other reportable incidents	47265	Dual-chamber implantable pacemaker, rate-responsive	CT1685	Cardiac pacemakers	E0717	Dyspnea	A1205	Loose or Intermittent Connection		
							F08	Hospitalization or Prolonged Hospitalization	A1205	Loose or Intermittent Connection		
							F12	Serious Injury/ Illness/ Impairment	A1205	Loose or Intermittent Connection		
							F1905	Device Revision or Replacement	A1205	Loose or Intermittent Connection		
							F23	Unexpected Medical Intervention	A1205	Loose or Intermittent Connection		
					CT257	Implantable cardiac pacemakers	E0717	Dyspnea	A1205	Loose or Intermittent Connection		
							F08	Hospitalization or Prolonged Hospitalization	A1205	Loose or Intermittent Connection		
							F12	Serious Injury/ Illness/ Impairment	A1205	Loose or Intermittent Connection		
							F1905	Device Revision or Replacement	A1205	Loose or Intermittent Connection		
							F23	Unexpected Medical Intervention	A1205	Loose or Intermittent Connection		
					CT2866	Implantable cardiac electrical stimulators	E0717	Dyspnea	A1205	Loose or Intermittent Connection		
							F08	Hospitalization or Prolonged Hospitalization	A1205	Loose or Intermittent Connection		
							F12	Serious Injury/ Illness/ Impairment	A1205	Loose or Intermittent Connection		
							F1905	Device Revision or Replacement	A1205	Loose or Intermittent Connection		
							F23	Unexpected Medical Intervention	A1205	Loose or Intermittent Connection		
15120	2024	All other reportable incidents	47205	Cardiac pulse generator programmer	CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A150105	Key or Button Unresponsive/not Working		
							F26	No Health Consequences or Impact	A150105	Key or Button Unresponsive/not Working		
15121	2024	All other reportable incidents	47270	Cardiac resynchronizatlon therapy implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	29/01/2025	29/01/2025
									A0908	Incorrect, Inadequate or Imprecise Result or Readings	29/01/2025	29/01/2025
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	29/01/2025	29/01/2025
									A0908	Incorrect, Inadequate or Imprecise Result or Readings	29/01/2025	29/01/2025
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	29/01/2025	29/01/2025
									A0908	Incorrect, Inadequate or Imprecise Result or Readings	29/01/2025	29/01/2025
15122	2024	All other reportable incidents	47265	Dual-chamber implantable pacemaker, rate-responsive	CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
							F24	Insufficient Health Impact Information	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
							F24	Insufficient Health Impact Information	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
							F24	Insufficient Health Impact Information	A0705	Battery Problem		

									A070504	Premature Discharge of Battery		
15123	2024	All other reportable incidents	47270	Cardiac resynchronization therapy implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A070909	Oversensing		
									A071302	Inappropriate/Inadequate Shock/Stimulation		
							F12	Serious Injury/ Illness/ Impairment	A070909	Oversensing		
									A071302	Inappropriate/Inadequate Shock/Stimulation		
15124	2024	All other reportable incidents	17882	Physiologic-monitoring defibrillation system	CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A0719	Unexpected Shutdown		
							F26	No Health Consequences or Impact	A0719	Unexpected Shutdown		
15125	2024	All other reportable incidents	47265	Dual-chamber implantable pacemaker, rate-responsive	CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A160601	Premature Elective Replacement Indicator	09/04/2024	09/04/2024
							F1905	Device Revision or Replacement	A160601	Premature Elective Replacement Indicator	09/04/2024	09/04/2024
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A160601	Premature Elective Replacement Indicator	09/04/2024	09/04/2024
							F1905	Device Revision or Replacement	A160601	Premature Elective Replacement Indicator	09/04/2024	09/04/2024
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A160601	Premature Elective Replacement Indicator	09/04/2024	09/04/2024
							F1905	Device Revision or Replacement	A160601	Premature Elective Replacement Indicator	09/04/2024	09/04/2024
15126	2024		17882	Physiologic-monitoring defibrillation system	CT1684	Cardiac pacemakers and associated devices	F26	No Health Consequences or Impact	A110601	Failure to Back-Up		
15127	2024	All other reportable incidents			CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A070101	Failure to Capture		
							F26	No Health Consequences or Impact	A070101	Failure to Capture		
15128	2024	All other reportable incidents	47205	Cardiac pulse generator programmer	CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A0203	Defective Device		
									A150105	Key or Button Unresponsive/not Working		
							F27	Problem Identified During Non-Clinical Procedure	A0203	Defective Device		
									A150105	Key or Button Unresponsive/not Working		
15129	2024	All other reportable incidents	47265	Dual-chamber implantable pacemaker, rate-responsive	CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem	21/03/2024	21/03/2024
									A070504	Premature Discharge of Battery	21/03/2024	21/03/2024
							F08	Hospitalization or Prolonged Hospitalization	A0705	Battery Problem	21/03/2024	21/03/2024
									A070504	Premature Discharge of Battery	21/03/2024	21/03/2024
							F1905	Device Revision or Replacement	A0705	Battery Problem	21/03/2024	21/03/2024
									A070504	Premature Discharge of Battery	21/03/2024	21/03/2024
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem	21/03/2024	21/03/2024
									A070504	Premature Discharge of Battery	21/03/2024	21/03/2024
							F08	Hospitalization or Prolonged Hospitalization	A0705	Battery Problem	21/03/2024	21/03/2024
									A070504	Premature Discharge of Battery	21/03/2024	21/03/2024
							F1905	Device Revision or Replacement	A0705	Battery Problem	21/03/2024	21/03/2024
									A070504	Premature Discharge of Battery	21/03/2024	21/03/2024
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem	21/03/2024	21/03/2024
									A070504	Premature Discharge of Battery	21/03/2024	21/03/2024
							F08	Hospitalization or Prolonged Hospitalization	A0705	Battery Problem	21/03/2024	21/03/2024
									A070504	Premature Discharge of Battery	21/03/2024	21/03/2024
							F1905	Device Revision or Replacement	A0705	Battery Problem	21/03/2024	21/03/2024
									A070504	Premature Discharge of Battery	21/03/2024	21/03/2024
15130	2024	All other reportable incidents	47270	Cardiac resynchronization therapy implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E2311	Discomfort	A010402	Migration		
							F26	No Health Consequences or Impact	A010402	Migration		
15131	2024	All other reportable incidents	47205	Cardiac pulse generator programmer	CT1684	Cardiac pacemakers and associated devices	E011903	Syncope/Fainting	A1104	Computer Operating System Problem		
							F12	Serious Injury/ Illness/ Impairment	A1104	Computer Operating System Problem		
							F23	Unexpected Medical Intervention	A1104	Computer Operating System Problem		
15132	2024	All other reportable incidents	47270	Cardiac resynchronization therapy implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
15133	2024	All other reportable incidents	47265	Dual-chamber implantable pacemaker, rate-responsive	CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem	03/03/2025	03/03/2025
									A070504	Premature Discharge of Battery	03/03/2025	03/03/2025
							F08	Hospitalization or Prolonged Hospitalization	A0705	Battery Problem	03/03/2025	03/03/2025
									A070504	Premature Discharge of Battery	03/03/2025	03/03/2025
							F1905	Device Revision or Replacement	A0705	Battery Problem	03/03/2025	03/03/2025
									A070504	Premature Discharge of Battery	03/03/2025	03/03/2025
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem	03/03/2025	03/03/2025
									A070504	Premature Discharge of Battery	03/03/2025	03/03/2025
							F08	Hospitalization or Prolonged Hospitalization	A0705	Battery Problem	03/03/2025	03/03/2025
									A070504	Premature Discharge of Battery	03/03/2025	03/03/2025
							F1905	Device Revision or Replacement	A0705	Battery Problem	03/03/2025	03/03/2025
									A070504	Premature Discharge of Battery	03/03/2025	03/03/2025
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem	03/03/2025	03/03/2025
									A070504	Premature Discharge of Battery	03/03/2025	03/03/2025
							F08	Hospitalization or Prolonged Hospitalization	A0705	Battery Problem	03/03/2025	03/03/2025
									A070504	Premature Discharge of Battery	03/03/2025	03/03/2025
							F1905	Device Revision or Replacement	A0705	Battery Problem	03/03/2025	03/03/2025
									A070504	Premature Discharge of Battery	03/03/2025	03/03/2025
15134	2024	All other reportable incidents	35852	Single-chamber implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
15135	2024	All other reportable incidents	47265	Dual-chamber implantable pacemaker, rate-responsive	CT141	Electrodes/Leads and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem	21/03/2024	21/03/2024
									A070504	Premature Discharge of Battery	21/03/2024	21/03/2024
							F08	Hospitalization or Prolonged Hospitalization	A0705	Battery Problem	21/03/2024	21/03/2024
									A070504	Premature Discharge of Battery	21/03/2024	21/03/2024
							F1905	Device Revision or Replacement	A0705	Battery Problem	21/03/2024	21/03/2024
									A070504	Premature Discharge of Battery	21/03/2024	21/03/2024
					CT145	Cardiovascular devices	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem	21/03/2024	21/03/2024
									A070504	Premature Discharge of Battery	21/03/2024	21/03/2024
							F08	Hospitalization or Prolonged Hospitalization	A0705	Battery Problem	21/03/2024	21/03/2024
									A070504	Premature Discharge of Battery	21/03/2024	21/03/2024
							F1905	Device Revision or Replacement	A0705	Battery Problem	21/03/2024	21/03/2024
									A070504	Premature Discharge of Battery	21/03/2024	21/03/2024
					CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem	21/03/2024	21/03/2024
									A070504	Premature Discharge of Battery	21/03/2024	21/03/2024
							F08	Hospitalization or Prolonged Hospitalization	A0705	Battery Problem	21/03/2024	21/03/2024
									A070504	Premature Discharge of Battery	21/03/2024	21/03/2024
							F1905	Device Revision or Replacement	A0705	Battery Problem	21/03/2024	21/03/2024
									A070504	Premature Discharge of Battery	21/03/2024	21/03/2024
					CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem	21/03/2024	21/03/2024
									A070504	Premature Discharge of Battery	21/03/2024	21/03/2024
							F08	Hospitalization or Prolonged Hospitalization	A0705	Battery Problem	21/03/2024	21/03/2024
									A070504	Premature Discharge of Battery	21/03/2024	21/03/2024
							F1905	Device Revision or Replacement	A0705	Battery Problem	21/03/2024	21/03/2024

									A070504	Premature Discharge of Battery	21/03/2024	21/03/2024
					CT233	Surgical	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem	21/03/2024	21/03/2024
									A070504	Premature Discharge of Battery	21/03/2024	21/03/2024
							F08	Hospitalization or Prolonged Hospitalization	A0705	Battery Problem	21/03/2024	21/03/2024
									A070504	Premature Discharge of Battery	21/03/2024	21/03/2024
							F1905	Device Revision or Replacement	A0705	Battery Problem	21/03/2024	21/03/2024
									A070504	Premature Discharge of Battery	21/03/2024	21/03/2024
					CT2406	Implantable	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem	21/03/2024	21/03/2024
									A070504	Premature Discharge of Battery	21/03/2024	21/03/2024
							F08	Hospitalization or Prolonged Hospitalization	A0705	Battery Problem	21/03/2024	21/03/2024
									A070504	Premature Discharge of Battery	21/03/2024	21/03/2024
							F1905	Device Revision or Replacement	A0705	Battery Problem	21/03/2024	21/03/2024
									A070504	Premature Discharge of Battery	21/03/2024	21/03/2024
					CT2407	Active implantable	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem	21/03/2024	21/03/2024
									A070504	Premature Discharge of Battery	21/03/2024	21/03/2024
							F08	Hospitalization or Prolonged Hospitalization	A0705	Battery Problem	21/03/2024	21/03/2024
									A070504	Premature Discharge of Battery	21/03/2024	21/03/2024
							F1905	Device Revision or Replacement	A0705	Battery Problem	21/03/2024	21/03/2024
									A070504	Premature Discharge of Battery	21/03/2024	21/03/2024
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem	21/03/2024	21/03/2024
									A070504	Premature Discharge of Battery	21/03/2024	21/03/2024
							F08	Hospitalization or Prolonged Hospitalization	A0705	Battery Problem	21/03/2024	21/03/2024
									A070504	Premature Discharge of Battery	21/03/2024	21/03/2024
							F1905	Device Revision or Replacement	A0705	Battery Problem	21/03/2024	21/03/2024
									A070504	Premature Discharge of Battery	21/03/2024	21/03/2024
					CT2727	Active implant devices/ancillaries	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem	21/03/2024	21/03/2024
									A070504	Premature Discharge of Battery	21/03/2024	21/03/2024
							F08	Hospitalization or Prolonged Hospitalization	A0705	Battery Problem	21/03/2024	21/03/2024
									A070504	Premature Discharge of Battery	21/03/2024	21/03/2024
							F1905	Device Revision or Replacement	A0705	Battery Problem	21/03/2024	21/03/2024
									A070504	Premature Discharge of Battery	21/03/2024	21/03/2024
					CT2750	Implantable cardiac electrical stimulators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem	21/03/2024	21/03/2024
									A070504	Premature Discharge of Battery	21/03/2024	21/03/2024
							F08	Hospitalization or Prolonged Hospitalization	A0705	Battery Problem	21/03/2024	21/03/2024
									A070504	Premature Discharge of Battery	21/03/2024	21/03/2024
							F1905	Device Revision or Replacement	A0705	Battery Problem	21/03/2024	21/03/2024
									A070504	Premature Discharge of Battery	21/03/2024	21/03/2024
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem	21/03/2024	21/03/2024
									A070504	Premature Discharge of Battery	21/03/2024	21/03/2024
							F08	Hospitalization or Prolonged Hospitalization	A0705	Battery Problem	21/03/2024	21/03/2024
									A070504	Premature Discharge of Battery	21/03/2024	21/03/2024
							F1905	Device Revision or Replacement	A0705	Battery Problem	21/03/2024	21/03/2024
									A070504	Premature Discharge of Battery	21/03/2024	21/03/2024
					CT2985	Active	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem	21/03/2024	21/03/2024
									A070504	Premature Discharge of Battery	21/03/2024	21/03/2024
							F08	Hospitalization or Prolonged Hospitalization	A0705	Battery Problem	21/03/2024	21/03/2024
									A070504	Premature Discharge of Battery	21/03/2024	21/03/2024
							F1905	Device Revision or Replacement	A0705	Battery Problem	21/03/2024	21/03/2024
									A070504	Premature Discharge of Battery	21/03/2024	21/03/2024
					CT3001	Implantable metal	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem	21/03/2024	21/03/2024
									A070504	Premature Discharge of Battery	21/03/2024	21/03/2024
							F08	Hospitalization or Prolonged Hospitalization	A0705	Battery Problem	21/03/2024	21/03/2024
									A070504	Premature Discharge of Battery	21/03/2024	21/03/2024
							F1905	Device Revision or Replacement	A0705	Battery Problem	21/03/2024	21/03/2024
									A070504	Premature Discharge of Battery	21/03/2024	21/03/2024
					CT3002	Implantable synthetic polymer	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem	21/03/2024	21/03/2024
									A070504	Premature Discharge of Battery	21/03/2024	21/03/2024
							F08	Hospitalization or Prolonged Hospitalization	A0705	Battery Problem	21/03/2024	21/03/2024
									A070504	Premature Discharge of Battery	21/03/2024	21/03/2024
							F1905	Device Revision or Replacement	A0705	Battery Problem	21/03/2024	21/03/2024
									A070504	Premature Discharge of Battery	21/03/2024	21/03/2024
					CT3006	Implantable inorganic material	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem	21/03/2024	21/03/2024
									A070504	Premature Discharge of Battery	21/03/2024	21/03/2024
							F08	Hospitalization or Prolonged Hospitalization	A0705	Battery Problem	21/03/2024	21/03/2024
									A070504	Premature Discharge of Battery	21/03/2024	21/03/2024
							F1905	Device Revision or Replacement	A0705	Battery Problem	21/03/2024	21/03/2024
									A070504	Premature Discharge of Battery	21/03/2024	21/03/2024
					CT314	Electrically powered	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem	21/03/2024	21/03/2024
									A070504	Premature Discharge of Battery	21/03/2024	21/03/2024
							F08	Hospitalization or Prolonged Hospitalization	A0705	Battery Problem	21/03/2024	21/03/2024
									A070504	Premature Discharge of Battery	21/03/2024	21/03/2024
							F1905	Device Revision or Replacement	A0705	Battery Problem	21/03/2024	21/03/2024
									A070504	Premature Discharge of Battery	21/03/2024	21/03/2024
					CT321	Long-term surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem	21/03/2024	21/03/2024
									A070504	Premature Discharge of Battery	21/03/2024	21/03/2024
							F08	Hospitalization or Prolonged Hospitalization	A0705	Battery Problem	21/03/2024	21/03/2024
									A070504	Premature Discharge of Battery	21/03/2024	21/03/2024
							F1905	Device Revision or Replacement	A0705	Battery Problem	21/03/2024	21/03/2024
									A070504	Premature Discharge of Battery	21/03/2024	21/03/2024
					CT449	Generators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem	21/03/2024	21/03/2024
									A070504	Premature Discharge of Battery	21/03/2024	21/03/2024
							F08	Hospitalization or Prolonged Hospitalization	A0705	Battery Problem	21/03/2024	21/03/2024
									A070504	Premature Discharge of Battery	21/03/2024	21/03/2024
							F1905	Device Revision or Replacement	A0705	Battery Problem	21/03/2024	21/03/2024
									A070504	Premature Discharge of Battery	21/03/2024	21/03/2024
					CT981	Single use	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem	21/03/2024	21/03/2024
									A070504	Premature Discharge of Battery	21/03/2024	21/03/2024
							F08	Hospitalization or Prolonged Hospitalization	A0705	Battery Problem	21/03/2024	21/03/2024

									A070504	Premature Discharge of Battery	21/03/2024	21/03/2024
							F1905	Device Revision or Replacement	A0705	Battery Problem	21/03/2024	21/03/2024
									A070504	Premature Discharge of Battery	21/03/2024	21/03/2024
					CT983	Surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem	21/03/2024	21/03/2024
									A070504	Premature Discharge of Battery	21/03/2024	21/03/2024
							F08	Hospitalization or Prolonged Hospitalization	A0705	Battery Problem	21/03/2024	21/03/2024
									A070504	Premature Discharge of Battery	21/03/2024	21/03/2024
							F1905	Device Revision or Replacement	A0705	Battery Problem	21/03/2024	21/03/2024
									A070504	Premature Discharge of Battery	21/03/2024	21/03/2024
15136	2024	All other reportable incidents	47265	Dual-chamber implantable pacemaker, rate-responsive	CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
15137	2024	Death	17882	Physiologic-monitoring defibrillation system	CT1684	Cardiac pacemakers and associated devices	E0623	Unspecified Heart Problem	A071301	Failure to Deliver Shock/Stimulation		
							F05	Delay to Treatment/ Therapy	A071301	Failure to Deliver Shock/Stimulation		
15138	2024		47265	Dual-chamber implantable pacemaker, rate-responsive	CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	18/03/2024	18/03/2024
							F1901	Additional Surgery	A24	No Apparent Device Problem	18/03/2024	18/03/2024
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	18/03/2024	18/03/2024
							F1901	Additional Surgery	A24	No Apparent Device Problem	18/03/2024	18/03/2024
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	18/03/2024	18/03/2024
							F1901	Additional Surgery	A24	No Apparent Device Problem	18/03/2024	18/03/2024
15139	2024	All other reportable incidents	47270	Cardiac resynchronization therapy implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E210401	Shock from Patient Lead(s)	A0712	Pacing Problem		
									A071302	Inappropriate/Inadequate Shock/Stimulation		
							F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A0712	Pacing Problem		
									A071302	Inappropriate/Inadequate Shock/Stimulation		
15140	2024	All other reportable incidents	35852	Single-chamber implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A0901	Audible Prompt/Feedback Problem		
							F27	Problem Identified During Non-Clinical Procedure	A0901	Audible Prompt/Feedback Problem		
15141	2024	All other reportable incidents	47267	Single-chamber implantable pacemaker, rate-responsive	CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	05/12/2023	05/12/2023
									A071102	Failure to Interrogate	05/12/2023	05/12/2023
									A071204	No Pacing	05/12/2023	05/12/2023
							F1901	Additional Surgery	A070504	Premature Discharge of Battery	05/12/2023	05/12/2023
									A071102	Failure to Interrogate	05/12/2023	05/12/2023
									A071204	No Pacing	05/12/2023	05/12/2023
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	05/12/2023	05/12/2023
									A071102	Failure to Interrogate	05/12/2023	05/12/2023
									A071204	No Pacing	05/12/2023	05/12/2023
							F1901	Additional Surgery	A070504	Premature Discharge of Battery	05/12/2023	05/12/2023
									A071102	Failure to Interrogate	05/12/2023	05/12/2023
									A071204	No Pacing	05/12/2023	05/12/2023
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	05/12/2023	05/12/2023
									A071102	Failure to Interrogate	05/12/2023	05/12/2023
									A071204	No Pacing	05/12/2023	05/12/2023
							F1901	Additional Surgery	A070504	Premature Discharge of Battery	05/12/2023	05/12/2023
									A071102	Failure to Interrogate	05/12/2023	05/12/2023
									A071204	No Pacing	05/12/2023	05/12/2023
15142	2024	All other reportable incidents	47265	Dual-chamber implantable pacemaker, rate-responsive	CT1685	Cardiac pacemakers	E0601	Arrhythmia	A071102	Failure to Interrogate	19/04/2024	19/04/2024
									A071204	No Pacing	19/04/2024	19/04/2024
							F1901	Additional Surgery	A071102	Failure to Interrogate	19/04/2024	19/04/2024
									A071204	No Pacing	19/04/2024	19/04/2024
					CT257	Implantable cardiac pacemakers	E0601	Arrhythmia	A071102	Failure to Interrogate	19/04/2024	19/04/2024
									A071204	No Pacing	19/04/2024	19/04/2024
							F1901	Additional Surgery	A071102	Failure to Interrogate	19/04/2024	19/04/2024
									A071204	No Pacing	19/04/2024	19/04/2024
					CT2866	Implantable cardiac electrical stimulators	E0601	Arrhythmia	A071102	Failure to Interrogate	19/04/2024	19/04/2024
									A071204	No Pacing	19/04/2024	19/04/2024
							F1901	Additional Surgery	A071102	Failure to Interrogate	19/04/2024	19/04/2024
									A071204	No Pacing	19/04/2024	19/04/2024
15143	2024	All other reportable incidents	47265	Dual-chamber implantable pacemaker, rate-responsive	CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	15/04/2024	15/04/2024
							F1901	Additional Surgery	A070504	Premature Discharge of Battery	15/04/2024	15/04/2024
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	15/04/2024	15/04/2024
							F1901	Additional Surgery	A070504	Premature Discharge of Battery	15/04/2024	15/04/2024
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	15/04/2024	15/04/2024
							F1901	Additional Surgery	A070504	Premature Discharge of Battery	15/04/2024	15/04/2024
15144	2024	All other reportable incidents	47267	Single-chamber implantable pacemaker, rate-responsive	CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem		
							F1905	Device Revision or Replacement	A0705	Battery Problem		
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem		
							F1905	Device Revision or Replacement	A0705	Battery Problem		
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem		
							F1905	Device Revision or Replacement	A0705	Battery Problem		
15145	2024	All other reportable incidents	17882	Physiologic-monitoring defibrillation system	CT1684	Cardiac pacemakers and associated devices	E2401	Insufficient Information	A2303	Improper or Incorrect Procedure or Method		
							F05	Delay to Treatment/ Therapy	A2303	Improper or Incorrect Procedure or Method		
15146	2024	All other reportable incidents	47265	Dual-chamber implantable pacemaker, rate-responsive	CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A09	Output Problem		
									A13	Communication or Transmission Problem		
							F26	No Health Consequences or Impact	A09	Output Problem		
									A13	Communication or Transmission Problem		
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A09	Output Problem		
									A13	Communication or Transmission Problem		
							F26	No Health Consequences or Impact	A09	Output Problem		
									A13	Communication or Transmission Problem		
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A09	Output Problem		
									A13	Communication or Transmission Problem		
							F26	No Health Consequences or Impact	A09	Output Problem		
									A13	Communication or Transmission Problem		
15147	2024	All other reportable incidents	47265	Dual-chamber implantable pacemaker, rate-responsive	CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A0908	Incorrect, Inadequate or Imprecise Result or Readings	02/02/2024	02/02/2024

							F08	Hospitalization or Prolonged Hospitalization	A0908	Incorrect, Inadequate or Imprecise Result or Readings	02/02/2024	02/02/2024
							F1905	Device Revision or Replacement	A0908	Incorrect, Inadequate or Imprecise Result or Readings	02/02/2024	02/02/2024
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A0908	Incorrect, Inadequate or Imprecise Result or Readings	02/02/2024	02/02/2024
							F08	Hospitalization or Prolonged Hospitalization	A0908	Incorrect, Inadequate or Imprecise Result or Readings	02/02/2024	02/02/2024
							F1905	Device Revision or Replacement	A0908	Incorrect, Inadequate or Imprecise Result or Readings	02/02/2024	02/02/2024
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A0908	Incorrect, Inadequate or Imprecise Result or Readings	02/02/2024	02/02/2024
							F08	Hospitalization or Prolonged Hospitalization	A0908	Incorrect, Inadequate or Imprecise Result or Readings	02/02/2024	02/02/2024
							F1905	Device Revision or Replacement	A0908	Incorrect, Inadequate or Imprecise Result or Readings	02/02/2024	02/02/2024
15148	2024		47265	Dual-chamber implantable pacemaker, rate-responsive	CT1685	Cardiac pacemakers	E060106	Heart Block	A0202	Defective Component		
							F05	Delay to Treatment/ Therapy	A0202	Defective Component		
					CT257	Implantable cardiac pacemakers	E060106	Heart Block	A0202	Defective Component		
							F05	Delay to Treatment/ Therapy	A0202	Defective Component		
					CT2866	Implantable cardiac electrical stimulators	E060106	Heart Block	A0202	Defective Component		
							F05	Delay to Treatment/ Therapy	A0202	Defective Component		
15149	2024	All other reportable incidents	35852	Single-chamber implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E0514	Thrombosis/Thrombus	A24	No Apparent Device Problem		
							F12	Serious Injury/ Illness/ Impairment	A24	No Apparent Device Problem		
15150	2024	All other reportable incidents			CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	12/06/2024	12/06/2024
									A0908	Incorrect, Inadequate or Imprecise Result or Readings	12/06/2024	12/06/2024
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	12/06/2024	12/06/2024
									A0908	Incorrect, Inadequate or Imprecise Result or Readings	12/06/2024	12/06/2024
							F1903	Device Explantation	A070504	Premature Discharge of Battery	12/06/2024	12/06/2024
									A0908	Incorrect, Inadequate or Imprecise Result or Readings	12/06/2024	12/06/2024
15151	2024	All other reportable incidents	47270	Cardiac resynchronization therapy implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E011903	Syncope/Fainting	A070909	Oversensing		
									A090801	Signal Artifact/Noise		
							F12	Serious Injury/ Illness/ Impairment	A070909	Oversensing		
									A090801	Signal Artifact/Noise		
15152	2024	All other reportable incidents	47265	Dual-chamber implantable pacemaker, rate-responsive	CT1685	Cardiac pacemakers	E0601	Arrhythmia	A070909	Oversensing		
									A071204	No Pacing		
							F12	Serious Injury/ Illness/ Impairment	A070909	Oversensing		
									A071204	No Pacing		
					CT257	Implantable cardiac pacemakers	E0601	Arrhythmia	A070909	Oversensing		
									A071204	No Pacing		
							F12	Serious Injury/ Illness/ Impairment	A070909	Oversensing		
									A071204	No Pacing		
					CT2866	Implantable cardiac electrical stimulators	E0601	Arrhythmia	A070909	Oversensing		
									A071204	No Pacing		
							F12	Serious Injury/ Illness/ Impairment	A070909	Oversensing		
									A071204	No Pacing		
15153	2024	All other reportable incidents	47265	Dual-chamber implantable pacemaker, rate-responsive	CT141	Electrodes/Leads and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A05	Mechanical Problem	28/04/2024	28/04/2024
							F26	No Health Consequences or Impact	A05	Mechanical Problem	28/04/2024	28/04/2024
					CT145	Cardiovascular devices	E2403	No Clinical Signs, Symptoms or Conditions	A05	Mechanical Problem	28/04/2024	28/04/2024
							F26	No Health Consequences or Impact	A05	Mechanical Problem	28/04/2024	28/04/2024
					CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A05	Mechanical Problem	28/04/2024	28/04/2024
							F26	No Health Consequences or Impact	A05	Mechanical Problem	28/04/2024	28/04/2024
					CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A05	Mechanical Problem	28/04/2024	28/04/2024
							F26	No Health Consequences or Impact	A05	Mechanical Problem	28/04/2024	28/04/2024
					CT233	Surgical	E2403	No Clinical Signs, Symptoms or Conditions	A05	Mechanical Problem	28/04/2024	28/04/2024
							F26	No Health Consequences or Impact	A05	Mechanical Problem	28/04/2024	28/04/2024
					CT2406	Implantable	E2403	No Clinical Signs, Symptoms or Conditions	A05	Mechanical Problem	28/04/2024	28/04/2024
							F26	No Health Consequences or Impact	A05	Mechanical Problem	28/04/2024	28/04/2024
					CT2407	Active implantable	E2403	No Clinical Signs, Symptoms or Conditions	A05	Mechanical Problem	28/04/2024	28/04/2024
							F26	No Health Consequences or Impact	A05	Mechanical Problem	28/04/2024	28/04/2024
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A05	Mechanical Problem	28/04/2024	28/04/2024
							F26	No Health Consequences or Impact	A05	Mechanical Problem	28/04/2024	28/04/2024
					CT2727	Active implant devices/ancillaries	E2403	No Clinical Signs, Symptoms or Conditions	A05	Mechanical Problem	28/04/2024	28/04/2024
							F26	No Health Consequences or Impact	A05	Mechanical Problem	28/04/2024	28/04/2024
					CT2750	Implantable cardiac electrical stimulators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A05	Mechanical Problem	28/04/2024	28/04/2024
							F26	No Health Consequences or Impact	A05	Mechanical Problem	28/04/2024	28/04/2024
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A05	Mechanical Problem	28/04/2024	28/04/2024
							F26	No Health Consequences or Impact	A05	Mechanical Problem	28/04/2024	28/04/2024
					CT2985	Active	E2403	No Clinical Signs, Symptoms or Conditions	A05	Mechanical Problem	28/04/2024	28/04/2024
							F26	No Health Consequences or Impact	A05	Mechanical Problem	28/04/2024	28/04/2024
					CT3001	Implantable metal	E2403	No Clinical Signs, Symptoms or Conditions	A05	Mechanical Problem	28/04/2024	28/04/2024
							F26	No Health Consequences or Impact	A05	Mechanical Problem	28/04/2024	28/04/2024
					CT3002	Implantable synthetic polymer	E2403	No Clinical Signs, Symptoms or Conditions	A05	Mechanical Problem	28/04/2024	28/04/2024
							F26	No Health Consequences or Impact	A05	Mechanical Problem	28/04/2024	28/04/2024
					CT3006	Implantable inorganic material	E2403	No Clinical Signs, Symptoms or Conditions	A05	Mechanical Problem	28/04/2024	28/04/2024
							F26	No Health Consequences or Impact	A05	Mechanical Problem	28/04/2024	28/04/2024
					CT314	Electrically powered	E2403	No Clinical Signs, Symptoms or Conditions	A05	Mechanical Problem	28/04/2024	28/04/2024
							F26	No Health Consequences or Impact	A05	Mechanical Problem	28/04/2024	28/04/2024
					CT321	Long-term surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A05	Mechanical Problem	28/04/2024	28/04/2024
							F26	No Health Consequences or Impact	A05	Mechanical Problem	28/04/2024	28/04/2024
					CT449	Generators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A05	Mechanical Problem	28/04/2024	28/04/2024
							F26	No Health Consequences or Impact	A05	Mechanical Problem	28/04/2024	28/04/2024
					CT981	Single use	E2403	No Clinical Signs, Symptoms or Conditions	A05	Mechanical Problem	28/04/2024	28/04/2024
							F26	No Health Consequences or Impact	A05	Mechanical Problem	28/04/2024	28/04/2024
					CT983	Surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A05	Mechanical Problem	28/04/2024	28/04/2024
							F26	No Health Consequences or Impact	A05	Mechanical Problem	28/04/2024	28/04/2024
15154	2024	All other reportable incidents	47270	Cardiac resynchronization therapy implantable	CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	07/05/2025	07/05/2025

				defibrillator					A0908	Incorrect, Inadequate or Imprecise Result or Readings	07/05/2025	07/05/2025
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	07/05/2025	07/05/2025
									A0908	Incorrect, Inadequate or Imprecise Result or Readings	07/05/2025	07/05/2025
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	07/05/2025	07/05/2025
									A0908	Incorrect, Inadequate or Imprecise Result or Readings	07/05/2025	07/05/2025
15155	2024	All other reportable incidents	35852	Single-chamber implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
									A0908	Incorrect, Inadequate or Imprecise Result or Readings		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
									A0908	Incorrect, Inadequate or Imprecise Result or Readings		
15156	2024	All other reportable incidents	47265	Dual-chamber implantable pacemaker, rate-responsive	CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A0908	Incorrect, Inadequate or Imprecise Result or Readings	30/01/2024	30/01/2024
							F08	Hospitalization or Prolonged Hospitalization	A0908	Incorrect, Inadequate or Imprecise Result or Readings	30/01/2024	30/01/2024
							F1905	Device Revision or Replacement	A0908	Incorrect, Inadequate or Imprecise Result or Readings	30/01/2024	30/01/2024
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A0908	Incorrect, Inadequate or Imprecise Result or Readings	30/01/2024	30/01/2024
							F08	Hospitalization or Prolonged Hospitalization	A0908	Incorrect, Inadequate or Imprecise Result or Readings	30/01/2024	30/01/2024
							F1905	Device Revision or Replacement	A0908	Incorrect, Inadequate or Imprecise Result or Readings	30/01/2024	30/01/2024
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A0908	Incorrect, Inadequate or Imprecise Result or Readings	30/01/2024	30/01/2024
							F08	Hospitalization or Prolonged Hospitalization	A0908	Incorrect, Inadequate or Imprecise Result or Readings	30/01/2024	30/01/2024
							F1905	Device Revision or Replacement	A0908	Incorrect, Inadequate or Imprecise Result or Readings	30/01/2024	30/01/2024
15157	2024	All other reportable incidents	35852	Single-chamber implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A072001	Electromagnetic Interference		
									A090801	Signal Artifact/Noise		
							F26	No Health Consequences or Impact	A072001	Electromagnetic Interference		
									A090801	Signal Artifact/Noise		
15158	2024	All other reportable incidents	47267	Single-chamber implantable pacemaker, rate-responsive	CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A09	Output Problem		
							F12	Serious Injury/ Illness/ Impairment	A09	Output Problem		
							F2307	Reprogramming of Device	A09	Output Problem		
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A09	Output Problem		
							F12	Serious Injury/ Illness/ Impairment	A09	Output Problem		
							F2307	Reprogramming of Device	A09	Output Problem		
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A09	Output Problem		
							F12	Serious Injury/ Illness/ Impairment	A09	Output Problem		
							F2307	Reprogramming of Device	A09	Output Problem		
15159	2024		47265	Dual-chamber implantable pacemaker, rate-responsive	CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	15/03/2024	15/03/2024
							F1901	Additional Surgery	A24	No Apparent Device Problem	15/03/2024	15/03/2024
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	15/03/2024	15/03/2024
							F1901	Additional Surgery	A24	No Apparent Device Problem	15/03/2024	15/03/2024
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	15/03/2024	15/03/2024
							F1901	Additional Surgery	A24	No Apparent Device Problem	15/03/2024	15/03/2024
15160	2024	All other reportable incidents	47265	Dual-chamber implantable pacemaker, rate-responsive	CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A0712	Pacing Problem		
							F26	No Health Consequences or Impact	A0712	Pacing Problem		
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A0712	Pacing Problem		
							F26	No Health Consequences or Impact	A0712	Pacing Problem		
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A0712	Pacing Problem		
							F26	No Health Consequences or Impact	A0712	Pacing Problem		
15161	2024	All other reportable incidents	47263	Cardiac resynchronization therapy implantable pacemaker	CT1685	Cardiac pacemakers	E020204	Malaise	A0712	Pacing Problem	30/04/2024	30/04/2024
									A0908	Incorrect, Inadequate or Imprecise Result or Readings	30/04/2024	30/04/2024
							E060101	Asystole	A0712	Pacing Problem	30/04/2024	30/04/2024
									A0908	Incorrect, Inadequate or Imprecise Result or Readings	30/04/2024	30/04/2024
							F08	Hospitalization or Prolonged Hospitalization	A0712	Pacing Problem	30/04/2024	30/04/2024
									A0908	Incorrect, Inadequate or Imprecise Result or Readings	30/04/2024	30/04/2024
							F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A0712	Pacing Problem	30/04/2024	30/04/2024
									A0908	Incorrect, Inadequate or Imprecise Result or Readings	30/04/2024	30/04/2024
							F1905	Device Revision or Replacement	A0712	Pacing Problem	30/04/2024	30/04/2024
									A0908	Incorrect, Inadequate or Imprecise Result or Readings	30/04/2024	30/04/2024
							F2301	Additional Device Required	A0712	Pacing Problem	30/04/2024	30/04/2024
									A0908	Incorrect, Inadequate or Imprecise Result or Readings	30/04/2024	30/04/2024
					CT257	Implantable cardiac pacemakers	E020204	Malaise	A0712	Pacing Problem	30/04/2024	30/04/2024
									A0908	Incorrect, Inadequate or Imprecise Result or Readings	30/04/2024	30/04/2024
							E060101	Asystole	A0712	Pacing Problem	30/04/2024	30/04/2024
									A0908	Incorrect, Inadequate or Imprecise Result or Readings	30/04/2024	30/04/2024
							F08	Hospitalization or Prolonged Hospitalization	A0712	Pacing Problem	30/04/2024	30/04/2024
									A0908	Incorrect, Inadequate or Imprecise Result or Readings	30/04/2024	30/04/2024
							F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A0712	Pacing Problem	30/04/2024	30/04/2024
									A0908	Incorrect, Inadequate or Imprecise Result or Readings	30/04/2024	30/04/2024
							F1905	Device Revision or Replacement	A0712	Pacing Problem	30/04/2024	30/04/2024
									A0908	Incorrect, Inadequate or Imprecise Result or Readings	30/04/2024	30/04/2024
							F2301	Additional Device Required	A0712	Pacing Problem	30/04/2024	30/04/2024
									A0908	Incorrect, Inadequate or Imprecise Result or Readings	30/04/2024	30/04/2024
					CT2866	Implantable cardiac electrical stimulators	E020204	Malaise	A0712	Pacing Problem	30/04/2024	30/04/2024

									A0908	Incorrect, Inadequate or Imprecise Result or Readings	30/04/2024	30/04/2024
							E060101	Asystole	A0712	Pacing Problem	30/04/2024	30/04/2024
									A0908	Incorrect, Inadequate or Imprecise Result or Readings	30/04/2024	30/04/2024
							F08	Hospitalization or Prolonged Hospitalization	A0712	Pacing Problem	30/04/2024	30/04/2024
									A0908	Incorrect, Inadequate or Imprecise Result or Readings	30/04/2024	30/04/2024
							F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A0712	Pacing Problem	30/04/2024	30/04/2024
									A0908	Incorrect, Inadequate or Imprecise Result or Readings	30/04/2024	30/04/2024
							F1905	Device Revision or Replacement	A0712	Pacing Problem	30/04/2024	30/04/2024
									A0908	Incorrect, Inadequate or Imprecise Result or Readings	30/04/2024	30/04/2024
							F2301	Additional Device Required	A0712	Pacing Problem	30/04/2024	30/04/2024
									A0908	Incorrect, Inadequate or Imprecise Result or Readings	30/04/2024	30/04/2024
15162	2024	All other reportable incidents	47265	Dual-chamber implantable pacemaker, rate-responsive	CT141	Electrodes/Leads and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A0908	Incorrect, Inadequate or Imprecise Result or Readings	29/04/2024	29/04/2024
							F08	Hospitalization or Prolonged Hospitalization	A0908	Incorrect, Inadequate or Imprecise Result or Readings	29/04/2024	29/04/2024
							F1905	Device Revision or Replacement	A0908	Incorrect, Inadequate or Imprecise Result or Readings	29/04/2024	29/04/2024
					CT145	Cardiovascular devices	E2403	No Clinical Signs, Symptoms or Conditions	A0908	Incorrect, Inadequate or Imprecise Result or Readings	29/04/2024	29/04/2024
							F08	Hospitalization or Prolonged Hospitalization	A0908	Incorrect, Inadequate or Imprecise Result or Readings	29/04/2024	29/04/2024
							F1905	Device Revision or Replacement	A0908	Incorrect, Inadequate or Imprecise Result or Readings	29/04/2024	29/04/2024
					CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A0908	Incorrect, Inadequate or Imprecise Result or Readings	29/04/2024	29/04/2024
							F08	Hospitalization or Prolonged Hospitalization	A0908	Incorrect, Inadequate or Imprecise Result or Readings	29/04/2024	29/04/2024
							F1905	Device Revision or Replacement	A0908	Incorrect, Inadequate or Imprecise Result or Readings	29/04/2024	29/04/2024
					CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A0908	Incorrect, Inadequate or Imprecise Result or Readings	29/04/2024	29/04/2024
							F08	Hospitalization or Prolonged Hospitalization	A0908	Incorrect, Inadequate or Imprecise Result or Readings	29/04/2024	29/04/2024
							F1905	Device Revision or Replacement	A0908	Incorrect, Inadequate or Imprecise Result or Readings	29/04/2024	29/04/2024
					CT233	Surgical	E2403	No Clinical Signs, Symptoms or Conditions	A0908	Incorrect, Inadequate or Imprecise Result or Readings	29/04/2024	29/04/2024
							F08	Hospitalization or Prolonged Hospitalization	A0908	Incorrect, Inadequate or Imprecise Result or Readings	29/04/2024	29/04/2024
							F1905	Device Revision or Replacement	A0908	Incorrect, Inadequate or Imprecise Result or Readings	29/04/2024	29/04/2024
					CT2406	Implantable	E2403	No Clinical Signs, Symptoms or Conditions	A0908	Incorrect, Inadequate or Imprecise Result or Readings	29/04/2024	29/04/2024
							F08	Hospitalization or Prolonged Hospitalization	A0908	Incorrect, Inadequate or Imprecise Result or Readings	29/04/2024	29/04/2024
							F1905	Device Revision or Replacement	A0908	Incorrect, Inadequate or Imprecise Result or Readings	29/04/2024	29/04/2024
					CT2407	Active implantable	E2403	No Clinical Signs, Symptoms or Conditions	A0908	Incorrect, Inadequate or Imprecise Result or Readings	29/04/2024	29/04/2024
							F08	Hospitalization or Prolonged Hospitalization	A0908	Incorrect, Inadequate or Imprecise Result or Readings	29/04/2024	29/04/2024
							F1905	Device Revision or Replacement	A0908	Incorrect, Inadequate or Imprecise Result or Readings	29/04/2024	29/04/2024
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A0908	Incorrect, Inadequate or Imprecise Result or Readings	29/04/2024	29/04/2024
							F08	Hospitalization or Prolonged Hospitalization	A0908	Incorrect, Inadequate or Imprecise Result or Readings	29/04/2024	29/04/2024
							F1905	Device Revision or Replacement	A0908	Incorrect, Inadequate or Imprecise Result or Readings	29/04/2024	29/04/2024
					CT2727	Active implant devices/ancillaries	E2403	No Clinical Signs, Symptoms or Conditions	A0908	Incorrect, Inadequate or Imprecise Result or Readings	29/04/2024	29/04/2024
							F08	Hospitalization or Prolonged Hospitalization	A0908	Incorrect, Inadequate or Imprecise Result or Readings	29/04/2024	29/04/2024
							F1905	Device Revision or Replacement	A0908	Incorrect, Inadequate or Imprecise Result or Readings	29/04/2024	29/04/2024
					CT2750	Implantable cardiac electrical stimulators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A0908	Incorrect, Inadequate or Imprecise Result or Readings	29/04/2024	29/04/2024
							F08	Hospitalization or Prolonged Hospitalization	A0908	Incorrect, Inadequate or Imprecise Result or Readings	29/04/2024	29/04/2024
							F1905	Device Revision or Replacement	A0908	Incorrect, Inadequate or Imprecise Result or Readings	29/04/2024	29/04/2024
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A0908	Incorrect, Inadequate or Imprecise Result or Readings	29/04/2024	29/04/2024
							F08	Hospitalization or Prolonged Hospitalization	A0908	Incorrect, Inadequate or Imprecise Result or Readings	29/04/2024	29/04/2024
							F1905	Device Revision or Replacement	A0908	Incorrect, Inadequate or Imprecise Result or Readings	29/04/2024	29/04/2024
					CT2985	Active	E2403	No Clinical Signs, Symptoms or Conditions	A0908	Incorrect, Inadequate or Imprecise Result or Readings	29/04/2024	29/04/2024
							F08	Hospitalization or Prolonged Hospitalization	A0908	Incorrect, Inadequate or Imprecise Result or Readings	29/04/2024	29/04/2024
							F1905	Device Revision or Replacement	A0908	Incorrect, Inadequate or Imprecise Result or Readings	29/04/2024	29/04/2024
					CT3001	Implantable metal	E2403	No Clinical Signs, Symptoms or Conditions	A0908	Incorrect, Inadequate or Imprecise Result or Readings	29/04/2024	29/04/2024
							F08	Hospitalization or Prolonged Hospitalization	A0908	Incorrect, Inadequate or Imprecise Result or Readings	29/04/2024	29/04/2024
							F1905	Device Revision or Replacement	A0908	Incorrect, Inadequate or Imprecise Result or Readings	29/04/2024	29/04/2024
					CT3002	Implantable synthetic polymer	E2403	No Clinical Signs, Symptoms or Conditions	A0908	Incorrect, Inadequate or Imprecise Result or Readings	29/04/2024	29/04/2024
							F08	Hospitalization or Prolonged Hospitalization	A0908	Incorrect, Inadequate or Imprecise Result or Readings	29/04/2024	29/04/2024

							F1905	Device Revision or Replacement	A0908	Incorrect, Inadequate or Imprecise Result or Readings	29/04/2024	29/04/2024
					CT3006	Implantable inorganic material	E2403	No Clinical Signs, Symptoms or Conditions	A0908	Incorrect, Inadequate or Imprecise Result or Readings	29/04/2024	29/04/2024
							F08	Hospitalization or Prolonged Hospitalization	A0908	Incorrect, Inadequate or Imprecise Result or Readings	29/04/2024	29/04/2024
							F1905	Device Revision or Replacement	A0908	Incorrect, Inadequate or Imprecise Result or Readings	29/04/2024	29/04/2024
					CT314	Electrically powered	E2403	No Clinical Signs, Symptoms or Conditions	A0908	Incorrect, Inadequate or Imprecise Result or Readings	29/04/2024	29/04/2024
							F08	Hospitalization or Prolonged Hospitalization	A0908	Incorrect, Inadequate or Imprecise Result or Readings	29/04/2024	29/04/2024
							F1905	Device Revision or Replacement	A0908	Incorrect, Inadequate or Imprecise Result or Readings	29/04/2024	29/04/2024
					CT321	Long-term surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A0908	Incorrect, Inadequate or Imprecise Result or Readings	29/04/2024	29/04/2024
							F08	Hospitalization or Prolonged Hospitalization	A0908	Incorrect, Inadequate or Imprecise Result or Readings	29/04/2024	29/04/2024
							F1905	Device Revision or Replacement	A0908	Incorrect, Inadequate or Imprecise Result or Readings	29/04/2024	29/04/2024
					CT449	Generators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A0908	Incorrect, Inadequate or Imprecise Result or Readings	29/04/2024	29/04/2024
							F08	Hospitalization or Prolonged Hospitalization	A0908	Incorrect, Inadequate or Imprecise Result or Readings	29/04/2024	29/04/2024
							F1905	Device Revision or Replacement	A0908	Incorrect, Inadequate or Imprecise Result or Readings	29/04/2024	29/04/2024
					CT981	Single use	E2403	No Clinical Signs, Symptoms or Conditions	A0908	Incorrect, Inadequate or Imprecise Result or Readings	29/04/2024	29/04/2024
							F08	Hospitalization or Prolonged Hospitalization	A0908	Incorrect, Inadequate or Imprecise Result or Readings	29/04/2024	29/04/2024
							F1905	Device Revision or Replacement	A0908	Incorrect, Inadequate or Imprecise Result or Readings	29/04/2024	29/04/2024
					CT983	Surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A0908	Incorrect, Inadequate or Imprecise Result or Readings	29/04/2024	29/04/2024
							F08	Hospitalization or Prolonged Hospitalization	A0908	Incorrect, Inadequate or Imprecise Result or Readings	29/04/2024	29/04/2024
							F1905	Device Revision or Replacement	A0908	Incorrect, Inadequate or Imprecise Result or Readings	29/04/2024	29/04/2024
15163	2024	All other reportable incidents	47265	Dual-chamber implantable pacemaker, rate-responsive	CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	29/04/2024	29/04/2024
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	29/04/2024	29/04/2024
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	29/04/2024	29/04/2024
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	29/04/2024	29/04/2024
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	29/04/2024	29/04/2024
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	29/04/2024	29/04/2024
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	29/04/2024	29/04/2024
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	29/04/2024	29/04/2024
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	29/04/2024	29/04/2024
15164	2024	All other reportable incidents	35852	Single-chamber implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A070101	Failure to Capture		
									A072201	High Impedance		
							F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A070101	Failure to Capture		
									A072201	High Impedance		
15165	2024	All other reportable incidents	35852	Single-chamber implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A0709	Device Sensing Problem		
							F12	Serious Injury/ Illness/ Impairment	A0709	Device Sensing Problem		
							F2307	Reprogramming of Device	A0709	Device Sensing Problem		
15166	2024	All other reportable incidents	47265	Dual-chamber implantable pacemaker, rate-responsive	CT1685	Cardiac pacemakers	E011903	Syncope/Fainting	A0908	Incorrect, Inadequate or Imprecise Result or Readings		
							F15	Recognized Device or Procedural Complication	A0908	Incorrect, Inadequate or Imprecise Result or Readings		
					CT257	Implantable cardiac pacemakers	E011903	Syncope/Fainting	A0908	Incorrect, Inadequate or Imprecise Result or Readings		
							F15	Recognized Device or Procedural Complication	A0908	Incorrect, Inadequate or Imprecise Result or Readings		
					CT2866	Implantable cardiac electrical stimulators	E011903	Syncope/Fainting	A0908	Incorrect, Inadequate or Imprecise Result or Readings		
							F15	Recognized Device or Procedural Complication	A0908	Incorrect, Inadequate or Imprecise Result or Readings		
15167	2024	All other reportable incidents	35852	Single-chamber implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E2104	Electric Shock	A0709	Device Sensing Problem	23/04/2024	23/04/2024
									A070909	Oversensing	23/04/2024	23/04/2024
									A071302	Inappropriate/Inadequate Shock/Stimulation	23/04/2024	23/04/2024
							F08	Hospitalization or Prolonged Hospitalization	A0709	Device Sensing Problem	23/04/2024	23/04/2024
									A070909	Oversensing	23/04/2024	23/04/2024
									A071302	Inappropriate/Inadequate Shock/Stimulation	23/04/2024	23/04/2024
							F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A0709	Device Sensing Problem	23/04/2024	23/04/2024
									A070909	Oversensing	23/04/2024	23/04/2024
									A071302	Inappropriate/Inadequate Shock/Stimulation	23/04/2024	23/04/2024
							F1903	Device Explantation	A0709	Device Sensing Problem	23/04/2024	23/04/2024
									A070909	Oversensing	23/04/2024	23/04/2024
									A071302	Inappropriate/Inadequate Shock/Stimulation	23/04/2024	23/04/2024
							F2203	Imaging Required	A0709	Device Sensing Problem	23/04/2024	23/04/2024
									A070909	Oversensing	23/04/2024	23/04/2024
									A071302	Inappropriate/Inadequate Shock/Stimulation	23/04/2024	23/04/2024
15168	2024	All other reportable incidents	17882	Physiologic-monitoring defibrillation system	CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A1501	Activation Problem		
							F27	Problem Identified During Non-Clinical Procedure	A1501	Activation Problem		
15169	2024	All other reportable incidents	47265	Dual-chamber implantable pacemaker, rate-responsive	CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A0705	Battery Problem		

									A070504	Premature Discharge of Battery		
15170	2024	All other reportable incidents	47270	Cardiac resynchronization therapy implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A05	Mechanical Problem	25/04/2024	25/04/2024
							F08	Hospitalization or Prolonged Hospitalization	A05	Mechanical Problem	25/04/2024	25/04/2024
							F12	Serious Injury/ Illness/ Impairment	A05	Mechanical Problem	25/04/2024	25/04/2024
							F1905	Device Revision or Replacement	A05	Mechanical Problem	25/04/2024	25/04/2024
15171	2024	All other reportable incidents	35852	Single-chamber implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	03/05/2024	03/05/2024
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	03/05/2024	03/05/2024
							F12	Serious Injury/ Illness/ Impairment	A070504	Premature Discharge of Battery	03/05/2024	03/05/2024
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	03/05/2024	03/05/2024
15172	2024	All other reportable incidents	35852	Single-chamber implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E210401	Shock from Patient Lead(s)	A071302	Inappropriate/Inadequate Shock/Stimulation		
							F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A071302	Inappropriate/Inadequate Shock/Stimulation		
15173	2024	All other reportable incidents	47265	Dual-chamber implantable pacemaker, rate-responsive	CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A0908	Incorrect, Inadequate or Imprecise Result or Readings	02/05/2024	02/05/2024
							F08	Hospitalization or Prolonged Hospitalization	A0908	Incorrect, Inadequate or Imprecise Result or Readings	02/05/2024	02/05/2024
							F1905	Device Revision or Replacement	A0908	Incorrect, Inadequate or Imprecise Result or Readings	02/05/2024	02/05/2024
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A0908	Incorrect, Inadequate or Imprecise Result or Readings	02/05/2024	02/05/2024
							F08	Hospitalization or Prolonged Hospitalization	A0908	Incorrect, Inadequate or Imprecise Result or Readings	02/05/2024	02/05/2024
							F1905	Device Revision or Replacement	A0908	Incorrect, Inadequate or Imprecise Result or Readings	02/05/2024	02/05/2024
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A0908	Incorrect, Inadequate or Imprecise Result or Readings	02/05/2024	02/05/2024
							F08	Hospitalization or Prolonged Hospitalization	A0908	Incorrect, Inadequate or Imprecise Result or Readings	02/05/2024	02/05/2024
							F1905	Device Revision or Replacement	A0908	Incorrect, Inadequate or Imprecise Result or Readings	02/05/2024	02/05/2024
15174	2024	All other reportable incidents	47265	Dual-chamber implantable pacemaker, rate-responsive	CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A0908	Incorrect, Inadequate or Imprecise Result or Readings	19/04/2024	19/04/2024
							F08	Hospitalization or Prolonged Hospitalization	A0908	Incorrect, Inadequate or Imprecise Result or Readings	19/04/2024	19/04/2024
							F1905	Device Revision or Replacement	A0908	Incorrect, Inadequate or Imprecise Result or Readings	19/04/2024	19/04/2024
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A0908	Incorrect, Inadequate or Imprecise Result or Readings	19/04/2024	19/04/2024
							F08	Hospitalization or Prolonged Hospitalization	A0908	Incorrect, Inadequate or Imprecise Result or Readings	19/04/2024	19/04/2024
							F1905	Device Revision or Replacement	A0908	Incorrect, Inadequate or Imprecise Result or Readings	19/04/2024	19/04/2024
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A0908	Incorrect, Inadequate or Imprecise Result or Readings	19/04/2024	19/04/2024
							F08	Hospitalization or Prolonged Hospitalization	A0908	Incorrect, Inadequate or Imprecise Result or Readings	19/04/2024	19/04/2024
							F1905	Device Revision or Replacement	A0908	Incorrect, Inadequate or Imprecise Result or Readings	19/04/2024	19/04/2024
15175	2024	All other reportable incidents			CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070910	Undersensing		
									A090804	False Positive Result		
							F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A070910	Undersensing		
									A090804	False Positive Result		
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A070910	Undersensing		
									A090804	False Positive Result		
							F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A070910	Undersensing		
									A090804	False Positive Result		
15176	2024	All other reportable incidents	47265	Dual-chamber implantable pacemaker, rate-responsive	CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
15177	2024	All other reportable incidents	47267	Single-chamber implantable pacemaker, rate-responsive	CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A072001	Electromagnetic Interference		
							F26	No Health Consequences or Impact	A072001	Electromagnetic Interference		
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A072001	Electromagnetic Interference		
							F26	No Health Consequences or Impact	A072001	Electromagnetic Interference		
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A072001	Electromagnetic Interference		
							F26	No Health Consequences or Impact	A072001	Electromagnetic Interference		
15178	2024	All other reportable incidents	35852	Single-chamber implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A0713	Defibrillation/Stimulation Problem		
							F12	Serious Injury/ Illness/ Impairment	A0713	Defibrillation/Stimulation Problem		
							F2307	Reprogramming of Device	A0713	Defibrillation/Stimulation Problem		
15179	2024		17882	Physiologic-monitoring defibrillation system	CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A0203	Defective Device		
									A1603	Failure of Device to Self-Test		
							F26	No Health Consequences or Impact	A0203	Defective Device		
									A1603	Failure of Device to Self-Test		
15180	2024		47265	Dual-chamber implantable pacemaker, rate-responsive	CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	29/04/2024	29/04/2024
							F1901	Additional Surgery	A24	No Apparent Device Problem	29/04/2024	29/04/2024
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	29/04/2024	29/04/2024
							F1901	Additional Surgery	A24	No Apparent Device Problem	29/04/2024	29/04/2024
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	29/04/2024	29/04/2024
							F1901	Additional Surgery	A24	No Apparent Device Problem	29/04/2024	29/04/2024
15181	2024	All other reportable incidents	17882	Physiologic-monitoring defibrillation system	CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070802	Intermittent Loss of Power		
							F27	Problem Identified During Non-Clinical Procedure	A070802	Intermittent Loss of Power		
15182	2024	All other reportable incidents	37265	Dual-chamber implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E210401	Shock from Patient Lead(s)	A071302	Inappropriate/Inadequate Shock/Stimulation		
							F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A071302	Inappropriate/Inadequate Shock/Stimulation		
15183	2024	All other reportable incidents	47265	Dual-chamber implantable pacemaker, rate-responsive	CT1685	Cardiac pacemakers	E2007	Fall	A0105	Lack of Effect	02/05/2024	02/05/2024
									A0705	Battery Problem	02/05/2024	02/05/2024
									A070504	Premature Discharge of Battery	02/05/2024	02/05/2024
									A070801	Complete Loss of Power	02/05/2024	02/05/2024
									A0711	Interrogation Problem	02/05/2024	02/05/2024
							E2403	No Clinical Signs, Symptoms or Conditions	A0105	Lack of Effect	02/05/2024	02/05/2024
									A0705	Battery Problem	02/05/2024	02/05/2024
									A070504	Premature Discharge of Battery	02/05/2024	02/05/2024

									A070801	Complete Loss of Power	02/05/2024	02/05/2024
									A0711	Interrogation Problem	02/05/2024	02/05/2024
							F05	Delay to Treatment/ Therapy	A0105	Lack of Effect	02/05/2024	02/05/2024
									A0705	Battery Problem	02/05/2024	02/05/2024
									A070504	Premature Discharge of Battery	02/05/2024	02/05/2024
									A070801	Complete Loss of Power	02/05/2024	02/05/2024
									A0711	Interrogation Problem	02/05/2024	02/05/2024
							F08	Hospitalization or Prolonged Hospitalization	A0105	Lack of Effect	02/05/2024	02/05/2024
									A0705	Battery Problem	02/05/2024	02/05/2024
									A070504	Premature Discharge of Battery	02/05/2024	02/05/2024
									A070801	Complete Loss of Power	02/05/2024	02/05/2024
									A0711	Interrogation Problem	02/05/2024	02/05/2024
							F1905	Device Revision or Replacement	A0105	Lack of Effect	02/05/2024	02/05/2024
									A0705	Battery Problem	02/05/2024	02/05/2024
									A070504	Premature Discharge of Battery	02/05/2024	02/05/2024
									A070801	Complete Loss of Power	02/05/2024	02/05/2024
									A0711	Interrogation Problem	02/05/2024	02/05/2024
					CT257	Implantable cardiac pacemakers	E2007	Fall	A0105	Lack of Effect	02/05/2024	02/05/2024
									A0705	Battery Problem	02/05/2024	02/05/2024
									A070504	Premature Discharge of Battery	02/05/2024	02/05/2024
									A070801	Complete Loss of Power	02/05/2024	02/05/2024
									A0711	Interrogation Problem	02/05/2024	02/05/2024
							E2403	No Clinical Signs, Symptoms or Conditions	A0105	Lack of Effect	02/05/2024	02/05/2024
									A0705	Battery Problem	02/05/2024	02/05/2024
									A070504	Premature Discharge of Battery	02/05/2024	02/05/2024
									A070801	Complete Loss of Power	02/05/2024	02/05/2024
									A0711	Interrogation Problem	02/05/2024	02/05/2024
							F05	Delay to Treatment/ Therapy	A0105	Lack of Effect	02/05/2024	02/05/2024
									A0705	Battery Problem	02/05/2024	02/05/2024
									A070504	Premature Discharge of Battery	02/05/2024	02/05/2024
									A070801	Complete Loss of Power	02/05/2024	02/05/2024
									A0711	Interrogation Problem	02/05/2024	02/05/2024
							F08	Hospitalization or Prolonged Hospitalization	A0105	Lack of Effect	02/05/2024	02/05/2024
									A0705	Battery Problem	02/05/2024	02/05/2024
									A070504	Premature Discharge of Battery	02/05/2024	02/05/2024
									A070801	Complete Loss of Power	02/05/2024	02/05/2024
									A0711	Interrogation Problem	02/05/2024	02/05/2024
							F1905	Device Revision or Replacement	A0105	Lack of Effect	02/05/2024	02/05/2024
									A0705	Battery Problem	02/05/2024	02/05/2024
									A070504	Premature Discharge of Battery	02/05/2024	02/05/2024
									A070801	Complete Loss of Power	02/05/2024	02/05/2024
									A0711	Interrogation Problem	02/05/2024	02/05/2024
					CT2866	Implantable cardiac electrical stimulators	E2007	Fall	A0105	Lack of Effect	02/05/2024	02/05/2024
									A0705	Battery Problem	02/05/2024	02/05/2024
									A070504	Premature Discharge of Battery	02/05/2024	02/05/2024
									A070801	Complete Loss of Power	02/05/2024	02/05/2024
									A0711	Interrogation Problem	02/05/2024	02/05/2024
							E2403	No Clinical Signs, Symptoms or Conditions	A0105	Lack of Effect	02/05/2024	02/05/2024
									A0705	Battery Problem	02/05/2024	02/05/2024
									A070504	Premature Discharge of Battery	02/05/2024	02/05/2024
									A070801	Complete Loss of Power	02/05/2024	02/05/2024
									A0711	Interrogation Problem	02/05/2024	02/05/2024
							F05	Delay to Treatment/ Therapy	A0105	Lack of Effect	02/05/2024	02/05/2024
									A0705	Battery Problem	02/05/2024	02/05/2024
									A070504	Premature Discharge of Battery	02/05/2024	02/05/2024
									A070801	Complete Loss of Power	02/05/2024	02/05/2024
									A0711	Interrogation Problem	02/05/2024	02/05/2024
							F08	Hospitalization or Prolonged Hospitalization	A0105	Lack of Effect	02/05/2024	02/05/2024
									A0705	Battery Problem	02/05/2024	02/05/2024
									A070504	Premature Discharge of Battery	02/05/2024	02/05/2024
									A070801	Complete Loss of Power	02/05/2024	02/05/2024
									A0711	Interrogation Problem	02/05/2024	02/05/2024
							F1905	Device Revision or Replacement	A0105	Lack of Effect	02/05/2024	02/05/2024
									A0705	Battery Problem	02/05/2024	02/05/2024
									A070504	Premature Discharge of Battery	02/05/2024	02/05/2024
									A070801	Complete Loss of Power	02/05/2024	02/05/2024
									A0711	Interrogation Problem	02/05/2024	02/05/2024
15184	2024		47263	Cardiac resynchronization therapy implantable pacemaker	CT1685	Cardiac pacemakers	E06	Heart	A24	No Apparent Device Problem		
							E0623	Unspecified Heart Problem	A24	No Apparent Device Problem		
							F26	No Health Consequences or Impact	A24	No Apparent Device Problem		
					CT257	Implantable cardiac pacemakers	E06	Heart	A24	No Apparent Device Problem		
							E0623	Unspecified Heart Problem	A24	No Apparent Device Problem		
							F26	No Health Consequences or Impact	A24	No Apparent Device Problem		
					CT2866	Implantable cardiac electrical stimulators	E06	Heart	A24	No Apparent Device Problem		
							E0623	Unspecified Heart Problem	A24	No Apparent Device Problem		
							F26	No Health Consequences or Impact	A24	No Apparent Device Problem		
15185	2024	All other reportable incidents	35852	Single-chamber implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E0112	Dizziness	A0709	Device Sensing Problem		
									A071302	Inappropriate/Inadequate Shock/Stimulation		
							E210401	Shock from Patient Lead(s)	A0709	Device Sensing Problem		
									A071302	Inappropriate/Inadequate Shock/Stimulation		
							E2310	Diaphoresis	A0709	Device Sensing Problem		
									A071302	Inappropriate/Inadequate Shock/Stimulation		
							F08	Hospitalization or Prolonged Hospitalization	A0709	Device Sensing Problem		
									A071302	Inappropriate/Inadequate Shock/Stimulation		
							F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A0709	Device Sensing Problem		
									A071302	Inappropriate/Inadequate Shock/Stimulation		
							F12	Serious Injury/ Illness/ Impairment	A0709	Device Sensing Problem		
									A071302	Inappropriate/Inadequate Shock/Stimulation		

						F1203	Life Threatening Illness or Injury	A0709	Device Sensing Problem		
								A071302	Inappropriate/Inadequate Shock/Stimulation		
						F2307	Reprogramming of Device	A0709	Device Sensing Problem		
								A071302	Inappropriate/Inadequate Shock/Stimulation		
15186	2024	All other reportable incidents	47263	Cardiac resynchronization therapy implantable pacemaker	CT1685	Cardiac pacemakers	E2330	Pain	A05	Mechanical Problem		
									A0712	Pacing Problem		
									A0722	Impedance Problem		
									A072201	High Impedance		
							F12	Serious Injury/ Illness/ Impairment	A05	Mechanical Problem		
									A0712	Pacing Problem		
									A0722	Impedance Problem		
									A072201	High Impedance		
							F2307	Reprogramming of Device	A05	Mechanical Problem		
									A0712	Pacing Problem		
									A0722	Impedance Problem		
									A072201	High Impedance		
					CT257	Implantable cardiac pacemakers	E2330	Pain	A05	Mechanical Problem		
									A0712	Pacing Problem		
									A0722	Impedance Problem		
									A072201	High Impedance		
							F12	Serious Injury/ Illness/ Impairment	A05	Mechanical Problem		
									A0712	Pacing Problem		
									A0722	Impedance Problem		
									A072201	High Impedance		
							F2307	Reprogramming of Device	A05	Mechanical Problem		
									A0712	Pacing Problem		
									A0722	Impedance Problem		
									A072201	High Impedance		
					CT2866	Implantable cardiac electrical stimulators	E2330	Pain	A05	Mechanical Problem		
									A0712	Pacing Problem		
									A0722	Impedance Problem		
									A072201	High Impedance		
							F12	Serious Injury/ Illness/ Impairment	A05	Mechanical Problem		
									A0712	Pacing Problem		
									A0722	Impedance Problem		
									A072201	High Impedance		
							F2307	Reprogramming of Device	A05	Mechanical Problem		
									A0712	Pacing Problem		
									A0722	Impedance Problem		
									A072201	High Impedance		
15187	2024	All other reportable incidents	47265	Dual-chamber implantable pacemaker, rate-responsive	CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A071102	Failure to Interrogate	17/05/2022	17/05/2022
									A071204	No Pacing	17/05/2022	17/05/2022
							F1901	Additional Surgery	A071102	Failure to Interrogate	17/05/2022	17/05/2022
									A071204	No Pacing	17/05/2022	17/05/2022
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A071102	Failure to Interrogate	17/05/2022	17/05/2022
									A071204	No Pacing	17/05/2022	17/05/2022
							F1901	Additional Surgery	A071102	Failure to Interrogate	17/05/2022	17/05/2022
									A071204	No Pacing	17/05/2022	17/05/2022
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A071102	Failure to Interrogate	17/05/2022	17/05/2022
									A071204	No Pacing	17/05/2022	17/05/2022
							F1901	Additional Surgery	A071102	Failure to Interrogate	17/05/2022	17/05/2022
									A071204	No Pacing	17/05/2022	17/05/2022
15188	2024	Death	17882	Physiologic-monitoring defibrillation system	CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A07	Electrical /Electronic Property Problem		
							F02	Death	A07	Electrical /Electronic Property Problem		
15189	2024	All other reportable incidents	35852	Single-chamber implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
									A0908	Incorrect, Inadequate or Imprecise Result or Readings		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
									A0908	Incorrect, Inadequate or Imprecise Result or Readings		
15190	2024	All other reportable incidents	35852	Single-chamber implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	07/05/2024	07/05/2024
									A071102	Failure to Interrogate	07/05/2024	07/05/2024
							F1901	Additional Surgery	A070504	Premature Discharge of Battery	07/05/2024	07/05/2024
									A071102	Failure to Interrogate	07/05/2024	07/05/2024
15191	2024		47265	Dual-chamber implantable pacemaker, rate-responsive	CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	30/04/2024	30/04/2024
							F1901	Additional Surgery	A24	No Apparent Device Problem	30/04/2024	30/04/2024
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	30/04/2024	30/04/2024
							F1901	Additional Surgery	A24	No Apparent Device Problem	30/04/2024	30/04/2024
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	30/04/2024	30/04/2024
							F1901	Additional Surgery	A24	No Apparent Device Problem	30/04/2024	30/04/2024
15192	2024		47265	Dual-chamber implantable pacemaker, rate-responsive	CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	05/12/2023	05/12/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	05/12/2023	05/12/2023
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	05/12/2023	05/12/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	05/12/2023	05/12/2023
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	05/12/2023	05/12/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	05/12/2023	05/12/2023
15193	2024		17882	Physiologic-monitoring defibrillation system	CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A0202	Defective Component		
							F26	No Health Consequences or Impact	A0202	Defective Component		
15194	2024	All other reportable incidents	47265	Dual-chamber implantable pacemaker, rate-responsive	CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A070909	Oversensing		
									A072001	Electromagnetic Interference		
									A090801	Signal Artifact/Noise		
							F26	No Health Consequences or Impact	A070909	Oversensing		
									A072001	Electromagnetic Interference		
									A090801	Signal Artifact/Noise		
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A070909	Oversensing		
									A072001	Electromagnetic Interference		
									A090801	Signal Artifact/Noise		
							F26	No Health Consequences or Impact	A070909	Oversensing		

									A072001	Electromagnetic Interference		
									A090801	Signal Artifact/Noise		
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A070909	Oversensing		
									A072001	Electromagnetic Interference		
									A090801	Signal Artifact/Noise		
							F26	No Health Consequences or Impact	A070909	Oversensing		
									A072001	Electromagnetic Interference		
									A090801	Signal Artifact/Noise		
15195	2024	All other reportable incidents	35852	Single-chamber implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E210401	Shock from Patient Lead(s)	A071302	Inappropriate/Inadequate Shock/Stimulation		
							F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A071302	Inappropriate/Inadequate Shock/Stimulation		
15196	2024	All other reportable incidents	47465	Cardiac contractility modulation system	CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A150105	Key or Button Unresponsive/not Working		
							F27	Problem Identified During Non-Clinical Procedure	A150105	Key or Button Unresponsive/not Working		
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A150105	Key or Button Unresponsive/not Working		
							F27	Problem Identified During Non-Clinical Procedure	A150105	Key or Button Unresponsive/not Working		
15197	2024	All other reportable incidents	47263	Cardiac resynchronization therapy implantable pacemaker	CT1684	Cardiac pacemakers and associated devices	E2103	Device Overstimulation of Tissue	A0908	Incorrect, Inadequate or Imprecise Result or Readings	23/04/2024	23/04/2024
							F08	Hospitalization or Prolonged Hospitalization	A0908	Incorrect, Inadequate or Imprecise Result or Readings	23/04/2024	23/04/2024
							F15	Recognized Device or Procedural Complication	A0908	Incorrect, Inadequate or Imprecise Result or Readings	23/04/2024	23/04/2024
							F1905	Device Revision or Replacement	A0908	Incorrect, Inadequate or Imprecise Result or Readings	23/04/2024	23/04/2024
					CT1685	Cardiac pacemakers	E2103	Device Overstimulation of Tissue	A0908	Incorrect, Inadequate or Imprecise Result or Readings	23/04/2024	23/04/2024
							F08	Hospitalization or Prolonged Hospitalization	A0908	Incorrect, Inadequate or Imprecise Result or Readings	23/04/2024	23/04/2024
							F15	Recognized Device or Procedural Complication	A0908	Incorrect, Inadequate or Imprecise Result or Readings	23/04/2024	23/04/2024
							F1905	Device Revision or Replacement	A0908	Incorrect, Inadequate or Imprecise Result or Readings	23/04/2024	23/04/2024
					CT257	Implantable cardiac pacemakers	E2103	Device Overstimulation of Tissue	A0908	Incorrect, Inadequate or Imprecise Result or Readings	23/04/2024	23/04/2024
							F08	Hospitalization or Prolonged Hospitalization	A0908	Incorrect, Inadequate or Imprecise Result or Readings	23/04/2024	23/04/2024
							F15	Recognized Device or Procedural Complication	A0908	Incorrect, Inadequate or Imprecise Result or Readings	23/04/2024	23/04/2024
							F1905	Device Revision or Replacement	A0908	Incorrect, Inadequate or Imprecise Result or Readings	23/04/2024	23/04/2024
					CT2866	Implantable cardiac electrical stimulators	E2103	Device Overstimulation of Tissue	A0908	Incorrect, Inadequate or Imprecise Result or Readings	23/04/2024	23/04/2024
							F08	Hospitalization or Prolonged Hospitalization	A0908	Incorrect, Inadequate or Imprecise Result or Readings	23/04/2024	23/04/2024
							F15	Recognized Device or Procedural Complication	A0908	Incorrect, Inadequate or Imprecise Result or Readings	23/04/2024	23/04/2024
							F1905	Device Revision or Replacement	A0908	Incorrect, Inadequate or Imprecise Result or Readings	23/04/2024	23/04/2024
15198	2024	All other reportable incidents	47265	Dual-chamber implantable pacemaker, rate-responsive	CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A070909	Oversensing		
									A090801	Signal Artifact/Noise		
							F26	No Health Consequences or Impact	A070909	Oversensing		
									A090801	Signal Artifact/Noise		
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A070909	Oversensing		
									A090801	Signal Artifact/Noise		
							F26	No Health Consequences or Impact	A070909	Oversensing		
									A090801	Signal Artifact/Noise		
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A070909	Oversensing		
									A090801	Signal Artifact/Noise		
							F26	No Health Consequences or Impact	A070909	Oversensing		
									A090801	Signal Artifact/Noise		
15199	2024	All other reportable incidents	17882	Physiologic-monitoring defibrillation system	CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A0719	Unexpected Shutdown		
							F27	Problem Identified During Non-Clinical Procedure	A0719	Unexpected Shutdown		
15200	2024	All other reportable incidents	47267	Single-chamber implantable pacemaker, rate-responsive	CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	19/04/2024	19/04/2024
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	19/04/2024	19/04/2024
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	19/04/2024	19/04/2024
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	19/04/2024	19/04/2024
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	19/04/2024	19/04/2024
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	19/04/2024	19/04/2024
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	19/04/2024	19/04/2024
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	19/04/2024	19/04/2024
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	19/04/2024	19/04/2024
15201	2024	All other reportable incidents	37265	Dual-chamber implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	14/04/2024	14/04/2024
							F1903	Device Explantation	A070504	Premature Discharge of Battery	14/04/2024	14/04/2024
15202	2024	All other reportable incidents	35852	Single-chamber implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E011903	Syncope/Fainting	A0709	Device Sensing Problem		
									A071301	Failure to Deliver Shock/Stimulation		
							F05	Delay to Treatment/ Therapy	A0709	Device Sensing Problem		
									A071301	Failure to Deliver Shock/Stimulation		
							F1001	Absence of Treatment	A0709	Device Sensing Problem		
									A071301	Failure to Deliver Shock/Stimulation		
							F1203	Life Threatening Illness or Injury	A0709	Device Sensing Problem		
									A071301	Failure to Deliver Shock/Stimulation		
							F2307	Reprogramming of Device	A0709	Device Sensing Problem		
									A071301	Failure to Deliver Shock/Stimulation		
15203	2024	All other reportable incidents	35852	Single-chamber implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A070909	Oversensing		
									A072001	Electromagnetic Interference		
									A090801	Signal Artifact/Noise		
							F26	No Health Consequences or Impact	A070909	Oversensing		
									A072001	Electromagnetic Interference		
									A090801	Signal Artifact/Noise		
15204	2024	All other reportable incidents	47267	Single-chamber implantable pacemaker, rate-responsive	CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem	14/02/2024	14/02/2024
									A070504	Premature Discharge of Battery	14/02/2024	14/02/2024
							F08	Hospitalization or Prolonged Hospitalization	A0705	Battery Problem	14/02/2024	14/02/2024
									A070504	Premature Discharge of Battery	14/02/2024	14/02/2024
							F1905	Device Revision or Replacement	A0705	Battery Problem	14/02/2024	14/02/2024

								A070504	Premature Discharge of Battery	14/02/2024	14/02/2024
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem	14/02/2024	14/02/2024
									A070504	Premature Discharge of Battery	14/02/2024	14/02/2024
							F08	Hospitalization or Prolonged Hospitalization	A0705	Battery Problem	14/02/2024	14/02/2024
									A070504	Premature Discharge of Battery	14/02/2024	14/02/2024
							F1905	Device Revision or Replacement	A0705	Battery Problem	14/02/2024	14/02/2024
									A070504	Premature Discharge of Battery	14/02/2024	14/02/2024
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem	14/02/2024	14/02/2024
									A070504	Premature Discharge of Battery	14/02/2024	14/02/2024
							F08	Hospitalization or Prolonged Hospitalization	A0705	Battery Problem	14/02/2024	14/02/2024
									A070504	Premature Discharge of Battery	14/02/2024	14/02/2024
							F1905	Device Revision or Replacement	A0705	Battery Problem	14/02/2024	14/02/2024
									A070504	Premature Discharge of Battery	14/02/2024	14/02/2024
15205	2024	All other reportable incidents	35852	Single-chamber implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E0119	Loss of Consciousness	A0712	Pacing Problem		
									A071202	Inaccurate Synchronization		
									A071302	Inappropriate/Inadequate Shock/Stimulation		
							E011903	Syncope/Fainting	A0712	Pacing Problem		
									A071202	Inaccurate Synchronization		
									A071302	Inappropriate/Inadequate Shock/Stimulation		
							E0601	Arrhythmia	A0712	Pacing Problem		
									A071202	Inaccurate Synchronization		
									A071302	Inappropriate/Inadequate Shock/Stimulation		
							E210401	Shock from Patient Lead(s)	A0712	Pacing Problem		
									A071202	Inaccurate Synchronization		
									A071302	Inappropriate/Inadequate Shock/Stimulation		
							F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A0712	Pacing Problem		
									A071202	Inaccurate Synchronization		
									A071302	Inappropriate/Inadequate Shock/Stimulation		
							F15	Recognized Device or Procedural Complication	A0712	Pacing Problem		
									A071202	Inaccurate Synchronization		
									A071302	Inappropriate/Inadequate Shock/Stimulation		
15206	2024	All other reportable incidents	47265	Dual-chamber implantable pacemaker, rate-responsive	CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
15207	2024	All other reportable incidents	47265	Dual-chamber implantable pacemaker, rate-responsive	CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A090811	Unable to Obtain Readings		
									A090812	Missing Test Results		
							F26	No Health Consequences or Impact	A090811	Unable to Obtain Readings		
									A090812	Missing Test Results		
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A090811	Unable to Obtain Readings		
									A090812	Missing Test Results		
							F26	No Health Consequences or Impact	A090811	Unable to Obtain Readings		
									A090812	Missing Test Results		
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A090811	Unable to Obtain Readings		
									A090812	Missing Test Results		
							F26	No Health Consequences or Impact	A090811	Unable to Obtain Readings		
									A090812	Missing Test Results		
15208	2024	All other reportable incidents	47265	Dual-chamber implantable pacemaker, rate-responsive	CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A070909	Oversensing		
							F26	No Health Consequences or Impact	A070909	Oversensing		
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A070909	Oversensing		
							F26	No Health Consequences or Impact	A070909	Oversensing		
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A070909	Oversensing		
							F26	No Health Consequences or Impact	A070909	Oversensing		
15209	2024		47205	Cardiac pulse generator programmer	CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A0203	Defective Device		
							F1905	Device Revision or Replacement	A0203	Defective Device		
15210	2024	All other reportable incidents	37265	Dual-chamber implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E210401	Shock from Patient Lead(s)	A0709	Device Sensing Problem		
									A071302	Inappropriate/Inadequate Shock/Stimulation		
							E2401	Insufficient Information	A0709	Device Sensing Problem		
									A071302	Inappropriate/Inadequate Shock/Stimulation		
							F24	Insufficient Health Impact Information	A0709	Device Sensing Problem		
									A071302	Inappropriate/Inadequate Shock/Stimulation		
15211	2024	All other reportable incidents	37265	Dual-chamber implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A070909	Oversensing		
									A072001	Electromagnetic Interference		
							F12	Serious Injury/ Illness/ Impairment	A070909	Oversensing		
									A072001	Electromagnetic Interference		
							F2307	Reprogramming of Device	A070909	Oversensing		
									A072001	Electromagnetic Interference		
15212	2024	All other reportable incidents	35852	Single-chamber implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E210401	Shock from Patient Lead(s)	A071302	Inappropriate/Inadequate Shock/Stimulation	10/05/2024	10/05/2024
									A1605	Reset Problem	10/05/2024	10/05/2024
							F08	Hospitalization or Prolonged Hospitalization	A071302	Inappropriate/Inadequate Shock/Stimulation	10/05/2024	10/05/2024
									A1605	Reset Problem	10/05/2024	10/05/2024
							F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A071302	Inappropriate/Inadequate Shock/Stimulation	10/05/2024	10/05/2024
									A1605	Reset Problem	10/05/2024	10/05/2024
							F1203	Life Threatening Illness or Injury	A071302	Inappropriate/Inadequate Shock/Stimulation	10/05/2024	10/05/2024
									A1605	Reset Problem	10/05/2024	10/05/2024
							F1903	Device Explantation	A071302	Inappropriate/Inadequate Shock/Stimulation	10/05/2024	10/05/2024
									A1605	Reset Problem	10/05/2024	10/05/2024
15213	2024	All other reportable incidents	47263	Cardiac resynchronization therapy implantable pacemaker	CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem	25/03/2024	25/03/2024
									A070504	Premature Discharge of Battery	25/03/2024	25/03/2024
							F08	Hospitalization or Prolonged Hospitalization	A0705	Battery Problem	25/03/2024	25/03/2024

									A070504	Premature Discharge of Battery	25/03/2024	25/03/2024
							F1905	Device Revision or Replacement	A0705	Battery Problem	25/03/2024	25/03/2024
									A070504	Premature Discharge of Battery	25/03/2024	25/03/2024
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem	25/03/2024	25/03/2024
									A070504	Premature Discharge of Battery	25/03/2024	25/03/2024
							F08	Hospitalization or Prolonged Hospitalization	A0705	Battery Problem	25/03/2024	25/03/2024
									A070504	Premature Discharge of Battery	25/03/2024	25/03/2024
							F1905	Device Revision or Replacement	A0705	Battery Problem	25/03/2024	25/03/2024
									A070504	Premature Discharge of Battery	25/03/2024	25/03/2024
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem	25/03/2024	25/03/2024
									A070504	Premature Discharge of Battery	25/03/2024	25/03/2024
							F08	Hospitalization or Prolonged Hospitalization	A0705	Battery Problem	25/03/2024	25/03/2024
									A070504	Premature Discharge of Battery	25/03/2024	25/03/2024
							F1905	Device Revision or Replacement	A0705	Battery Problem	25/03/2024	25/03/2024
									A070504	Premature Discharge of Battery	25/03/2024	25/03/2024
15214	2024	All other reportable incidents	47270	Cardiac resynchronization therapy implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A072001	Electromagnetic Interference		
							F26	No Health Consequences or Impact	A072001	Electromagnetic Interference		
15215	2024	All other reportable incidents	47270	Cardiac resynchronization therapy implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E0138	Undesired Nerve Stimulation	A0723	Unexpected Mode Switch	17/10/2023	17/10/2023
							F1905	Device Revision or Replacement	A0723	Unexpected Mode Switch	17/10/2023	17/10/2023
15216	2024	All other reportable incidents	37265	Dual-chamber implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E210401	Shock from Patient Lead(s)	A0709	Device Sensing Problem		
									A071302	Inappropriate/Inadequate Shock/Stimulation		
							F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A0709	Device Sensing Problem		
									A071302	Inappropriate/Inadequate Shock/Stimulation		
							F1203	Life Threatening Illness or Injury	A0709	Device Sensing Problem		
									A071302	Inappropriate/Inadequate Shock/Stimulation		
							F2307	Reprogramming of Device	A0709	Device Sensing Problem		
									A071302	Inappropriate/Inadequate Shock/Stimulation		
15217	2024	All other reportable incidents	47265	Dual-chamber implantable pacemaker, rate-responsive	CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem	13/05/2024	13/05/2024
									A070504	Premature Discharge of Battery	13/05/2024	13/05/2024
							F08	Hospitalization or Prolonged Hospitalization	A0705	Battery Problem	13/05/2024	13/05/2024
									A070504	Premature Discharge of Battery	13/05/2024	13/05/2024
							F1905	Device Revision or Replacement	A0705	Battery Problem	13/05/2024	13/05/2024
									A070504	Premature Discharge of Battery	13/05/2024	13/05/2024
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem	13/05/2024	13/05/2024
									A070504	Premature Discharge of Battery	13/05/2024	13/05/2024
							F08	Hospitalization or Prolonged Hospitalization	A0705	Battery Problem	13/05/2024	13/05/2024
									A070504	Premature Discharge of Battery	13/05/2024	13/05/2024
							F1905	Device Revision or Replacement	A0705	Battery Problem	13/05/2024	13/05/2024
									A070504	Premature Discharge of Battery	13/05/2024	13/05/2024
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem	13/05/2024	13/05/2024
									A070504	Premature Discharge of Battery	13/05/2024	13/05/2024
							F08	Hospitalization or Prolonged Hospitalization	A0705	Battery Problem	13/05/2024	13/05/2024
									A070504	Premature Discharge of Battery	13/05/2024	13/05/2024
							F1905	Device Revision or Replacement	A0705	Battery Problem	13/05/2024	13/05/2024
									A070504	Premature Discharge of Battery	13/05/2024	13/05/2024
15218	2024	All other reportable incidents	35852	Single-chamber implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	14/05/2024	14/05/2024
									A071102	Failure to Interrogate	14/05/2024	14/05/2024
							F1901	Additional Surgery	A070504	Premature Discharge of Battery	14/05/2024	14/05/2024
									A071102	Failure to Interrogate	14/05/2024	14/05/2024
15219	2024	All other reportable incidents	47205	Cardiac pulse generator programmer	CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A0203	Defective Device		
									A1102	Application Program Problem		
							F26	No Health Consequences or Impact	A0203	Defective Device		
									A1102	Application Program Problem		
15220	2024	All other reportable incidents	17882	Physiologic-monitoring defibrillation system	CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070912	Incorrect Interpretation of Signal		
							F27	Problem Identified During Non-Clinical Procedure	A070912	Incorrect Interpretation of Signal		
15221	2024	All other reportable incidents	17882	Physiologic-monitoring defibrillation system	CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070803	Failure to Power Up		
							F27	Problem Identified During Non-Clinical Procedure	A070803	Failure to Power Up		
15222	2024	All other reportable incidents	17882	Physiologic-monitoring defibrillation system	CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070803	Failure to Power Up		
							F27	Problem Identified During Non-Clinical Procedure	A070803	Failure to Power Up		
15223	2024	All other reportable incidents	47205	Cardiac pulse generator programmer	CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070803	Failure to Power Up		
									A150105	Key or Button Unresponsive/not Working		
							F27	Problem Identified During Non-Clinical Procedure	A070803	Failure to Power Up		
									A150105	Key or Button Unresponsive/not Working		
15224	2024	All other reportable incidents	17882	Physiologic-monitoring defibrillation system	CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A090402	Failure to Deliver Energy		
							F27	Problem Identified During Non-Clinical Procedure	A090402	Failure to Deliver Energy		
15225	2024	All other reportable incidents	17882	Physiologic-monitoring defibrillation system	CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A071204	No Pacing		
							F27	Problem Identified During Non-Clinical Procedure	A071204	No Pacing		
15226	2024	All other reportable incidents	47270	Cardiac resynchronization therapy implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A070101	Failure to Capture		
									A0709	Device Sensing Problem		
									A0722	Impedance Problem		
							F26	No Health Consequences or Impact	A070101	Failure to Capture		
									A0709	Device Sensing Problem		
									A0722	Impedance Problem		
15227	2024	All other reportable incidents	35852	Single-chamber implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A070909	Oversensing		
									A090801	Signal Artifact/Noise		
							F12	Serious Injury/ Illness/ Impairment	A070909	Oversensing		
									A090801	Signal Artifact/Noise		
							F2307	Reprogramming of Device	A070909	Oversensing		
									A090801	Signal Artifact/Noise		
15228	2024	Death	35852	Single-chamber implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E060110	Ventricular Fibrillation	A0709	Device Sensing Problem		
									A071301	Failure to Deliver Shock/Stimulation		
							F02	Death	A0709	Device Sensing Problem		
									A071301	Failure to Deliver Shock/Stimulation		
15229	2024	All other reportable incidents	47265	Dual-chamber implantable pacemaker, rate-responsive	CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem		

								A070504	Premature Discharge of Battery		
						F26	No Health Consequences or Impact	A0705	Battery Problem		
								A070504	Premature Discharge of Battery		
				CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem		
								A070504	Premature Discharge of Battery		
						F26	No Health Consequences or Impact	A0705	Battery Problem		
								A070504	Premature Discharge of Battery		
15230	2024	All other reportable incidents			CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem	25/06/2024	25/06/2024
									A070504	Premature Discharge of Battery	25/06/2024	25/06/2024
						F08	Hospitalization or Prolonged Hospitalization	A0705	Battery Problem	25/06/2024	25/06/2024
								A070504	Premature Discharge of Battery	25/06/2024	25/06/2024
						F1905	Device Revision or Replacement	A0705	Battery Problem	25/06/2024	25/06/2024
								A070504	Premature Discharge of Battery	25/06/2024	25/06/2024
15231	2024	All other reportable incidents	47265	Dual-chamber implantable pacemaker, rate-responsive	CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	02/05/2024	02/05/2024
									A0711	Interrogation Problem	02/05/2024	02/05/2024
						F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	02/05/2024	02/05/2024
								A0711	Interrogation Problem	02/05/2024	02/05/2024
						F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	02/05/2024	02/05/2024
								A0711	Interrogation Problem	02/05/2024	02/05/2024
				CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	02/05/2024	02/05/2024
								A0711	Interrogation Problem	02/05/2024	02/05/2024
						F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	02/05/2024	02/05/2024
								A0711	Interrogation Problem	02/05/2024	02/05/2024
						F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	02/05/2024	02/05/2024
								A0711	Interrogation Problem	02/05/2024	02/05/2024
				CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	02/05/2024	02/05/2024
								A0711	Interrogation Problem	02/05/2024	02/05/2024
						F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	02/05/2024	02/05/2024
								A0711	Interrogation Problem	02/05/2024	02/05/2024
						F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	02/05/2024	02/05/2024
								A0711	Interrogation Problem	02/05/2024	02/05/2024
15232	2024		47267	Single-chamber implantable pacemaker, rate-responsive	CT141	Electrodes/Leads and associated devices	E012207	Twitching	A0203	Defective Device		
							F1905	Device Revision or Replacement	A0203	Defective Device		
					CT145	Cardiovascular devices	E012207	Twitching	A0203	Defective Device		
							F1905	Device Revision or Replacement	A0203	Defective Device		
					CT1684	Cardiac pacemakers and associated devices	E012207	Twitching	A0203	Defective Device		
							F1905	Device Revision or Replacement	A0203	Defective Device		
					CT1685	Cardiac pacemakers	E012207	Twitching	A0203	Defective Device		
							F1905	Device Revision or Replacement	A0203	Defective Device		
					CT233	Surgical	E012207	Twitching	A0203	Defective Device		
							F1905	Device Revision or Replacement	A0203	Defective Device		
					CT2406	Implantable	E012207	Twitching	A0203	Defective Device		
							F1905	Device Revision or Replacement	A0203	Defective Device		
					CT2407	Active implantable	E012207	Twitching	A0203	Defective Device		
							F1905	Device Revision or Replacement	A0203	Defective Device		
					CT257	Implantable cardiac pacemakers	E012207	Twitching	A0203	Defective Device		
							F1905	Device Revision or Replacement	A0203	Defective Device		
					CT2727	Active implant devices/ancillaries	E012207	Twitching	A0203	Defective Device		
							F1905	Device Revision or Replacement	A0203	Defective Device		
					CT2750	Implantable cardiac electrical stimulators and associated devices	E012207	Twitching	A0203	Defective Device		
							F1905	Device Revision or Replacement	A0203	Defective Device		
					CT2866	Implantable cardiac electrical stimulators	E012207	Twitching	A0203	Defective Device		
							F1905	Device Revision or Replacement	A0203	Defective Device		
					CT2985	Active	E012207	Twitching	A0203	Defective Device		
							F1905	Device Revision or Replacement	A0203	Defective Device		
					CT3001	Implantable metal	E012207	Twitching	A0203	Defective Device		
							F1905	Device Revision or Replacement	A0203	Defective Device		
					CT3002	Implantable synthetic polymer	E012207	Twitching	A0203	Defective Device		
							F1905	Device Revision or Replacement	A0203	Defective Device		
					CT3006	Implantable inorganic material	E012207	Twitching	A0203	Defective Device		
							F1905	Device Revision or Replacement	A0203	Defective Device		
					CT314	Electrically powered	E012207	Twitching	A0203	Defective Device		
							F1905	Device Revision or Replacement	A0203	Defective Device		
					CT321	Long-term surgical invasive	E012207	Twitching	A0203	Defective Device		
							F1905	Device Revision or Replacement	A0203	Defective Device		
					CT449	Generators and associated devices	E012207	Twitching	A0203	Defective Device		
							F1905	Device Revision or Replacement	A0203	Defective Device		
					CT981	Single use	E012207	Twitching	A0203	Defective Device		
							F1905	Device Revision or Replacement	A0203	Defective Device		
					CT983	Surgical invasive	E012207	Twitching	A0203	Defective Device		
							F1905	Device Revision or Replacement	A0203	Defective Device		
15233	2024		35852	Single-chamber implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E0601	Arrhythmia	A070912	Incorrect Interpretation of Signal		
									A071301	Failure to Deliver Shock/Stimulation		
							F23	Unexpected Medical Intervention	A070912	Incorrect Interpretation of Signal		
									A071301	Failure to Deliver Shock/Stimulation		
15234	2024	All other reportable incidents	47265	Dual-chamber implantable pacemaker, rate-responsive	CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
15235	2024	All other reportable incidents	35224	External pacemaker, epicardial pacing	CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A0709	Device Sensing Problem		
									A09	Output Problem		
							F26	No Health Consequences or Impact	A0709	Device Sensing Problem		
									A09	Output Problem		
15236	2024	All other reportable incidents	35852	Single-chamber implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E0601	Arrhythmia	A070909	Oversensing		
									A0712	Pacing Problem		
							F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A070909	Oversensing		

								A0712	Pacing Problem			
15237	2024	All other reportable incidents	47263	Cardiac resynchronization therapy implantable pacemaker	CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A0908	Incorrect, Inadequate or Imprecise Result or Readings	14/05/2024	14/05/2024
							F08	Hospitalization or Prolonged Hospitalization	A0908	Incorrect, Inadequate or Imprecise Result or Readings	14/05/2024	14/05/2024
							F1905	Device Revision or Replacement	A0908	Incorrect, Inadequate or Imprecise Result or Readings	14/05/2024	14/05/2024
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A0908	Incorrect, Inadequate or Imprecise Result or Readings	14/05/2024	14/05/2024
							F08	Hospitalization or Prolonged Hospitalization	A0908	Incorrect, Inadequate or Imprecise Result or Readings	14/05/2024	14/05/2024
							F1905	Device Revision or Replacement	A0908	Incorrect, Inadequate or Imprecise Result or Readings	14/05/2024	14/05/2024
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A0908	Incorrect, Inadequate or Imprecise Result or Readings	14/05/2024	14/05/2024
							F08	Hospitalization or Prolonged Hospitalization	A0908	Incorrect, Inadequate or Imprecise Result or Readings	14/05/2024	14/05/2024
							F1905	Device Revision or Replacement	A0908	Incorrect, Inadequate or Imprecise Result or Readings	14/05/2024	14/05/2024
15238	2024	All other reportable incidents	35852	Single-chamber implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A070909	Oversensing		
									A071302	Inappropriate/Inadequate Shock/Stimulation		
									A090805	Incorrect Measurement		
							F12	Serious Injury/ Illness/ Impairment	A070909	Oversensing		
									A071302	Inappropriate/Inadequate Shock/Stimulation		
									A090805	Incorrect Measurement		
15239	2024	All other reportable incidents	47265	Dual-chamber implantable pacemaker, rate-responsive	CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A27	Appropriate Device Problem Term/Code not Available		
							F12	Serious Injury/ Illness/ Impairment	A27	Appropriate Device Problem Term/Code not Available		
							F2307	Reprogramming of Device	A27	Appropriate Device Problem Term/Code not Available		
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A27	Appropriate Device Problem Term/Code not Available		
							F12	Serious Injury/ Illness/ Impairment	A27	Appropriate Device Problem Term/Code not Available		
							F2307	Reprogramming of Device	A27	Appropriate Device Problem Term/Code not Available		
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A27	Appropriate Device Problem Term/Code not Available		
							F12	Serious Injury/ Illness/ Impairment	A27	Appropriate Device Problem Term/Code not Available		
							F2307	Reprogramming of Device	A27	Appropriate Device Problem Term/Code not Available		
15240	2024	All other reportable incidents	17882	Physiologic-monitoring defibrillation system	CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A0708	Power Problem		
							F27	Problem Identified During Non-Clinical Procedure	A0708	Power Problem		
15241	2024	All other reportable incidents	47265	Dual-chamber implantable pacemaker, rate-responsive	CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	02/09/2024	02/09/2024
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	02/09/2024	02/09/2024
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	02/09/2024	02/09/2024
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	02/09/2024	02/09/2024
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	02/09/2024	02/09/2024
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	02/09/2024	02/09/2024
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	02/09/2024	02/09/2024
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	02/09/2024	02/09/2024
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	02/09/2024	02/09/2024
15242	2024	All other reportable incidents	47265	Dual-chamber implantable pacemaker, rate-responsive	CT1685	Cardiac pacemakers	E0623	Unspecified Heart Problem	A0908	Incorrect, Inadequate or Imprecise Result or Readings	08/05/2024	08/05/2024
							F08	Hospitalization or Prolonged Hospitalization	A0908	Incorrect, Inadequate or Imprecise Result or Readings	08/05/2024	08/05/2024
							F1905	Device Revision or Replacement	A0908	Incorrect, Inadequate or Imprecise Result or Readings	08/05/2024	08/05/2024
					CT257	Implantable cardiac pacemakers	E0623	Unspecified Heart Problem	A0908	Incorrect, Inadequate or Imprecise Result or Readings	08/05/2024	08/05/2024
							F08	Hospitalization or Prolonged Hospitalization	A0908	Incorrect, Inadequate or Imprecise Result or Readings	08/05/2024	08/05/2024
							F1905	Device Revision or Replacement	A0908	Incorrect, Inadequate or Imprecise Result or Readings	08/05/2024	08/05/2024
					CT2866	Implantable cardiac electrical stimulators	E0623	Unspecified Heart Problem	A0908	Incorrect, Inadequate or Imprecise Result or Readings	08/05/2024	08/05/2024
							F08	Hospitalization or Prolonged Hospitalization	A0908	Incorrect, Inadequate or Imprecise Result or Readings	08/05/2024	08/05/2024
							F1905	Device Revision or Replacement	A0908	Incorrect, Inadequate or Imprecise Result or Readings	08/05/2024	08/05/2024
15243	2024	All other reportable incidents	47270	Cardiac resynchronization therapy implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E011903	Syncope/Fainting	A071201	Failure to Convert Rhythm		
							E2402	Appropriate Clinical Signs, Symptoms and Conditions Term/Code not Available	A071201	Failure to Convert Rhythm		
							F12	Serious Injury/ Illness/ Impairment	A071201	Failure to Convert Rhythm		
15244	2024	All other reportable incidents	37265	Dual-chamber implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A070909	Oversensing		
									A072201	High Impedance		
									A090801	Signal Artifact/Noise		
							F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A070909	Oversensing		
									A072201	High Impedance		
									A090801	Signal Artifact/Noise		
15245	2024	All other reportable incidents	47263	Cardiac resynchronization therapy implantable pacemaker	CT1685	Cardiac pacemakers	E011902	Presyncope	A070909	Oversensing	12/10/2023	12/10/2023
									A0712	Pacing Problem	12/10/2023	12/10/2023
									A072201	High Impedance	12/10/2023	12/10/2023
									A090801	Signal Artifact/Noise	12/10/2023	12/10/2023
									A12	Connection Problem	12/10/2023	12/10/2023
							E060101	Asystole	A070909	Oversensing	12/10/2023	12/10/2023
									A0712	Pacing Problem	12/10/2023	12/10/2023
									A072201	High Impedance	12/10/2023	12/10/2023
									A090801	Signal Artifact/Noise	12/10/2023	12/10/2023
									A12	Connection Problem	12/10/2023	12/10/2023
							F08	Hospitalization or Prolonged Hospitalization	A070909	Oversensing	12/10/2023	12/10/2023
									A0712	Pacing Problem	12/10/2023	12/10/2023

									A072201	High Impedance	12/10/2023	12/10/2023
									A090801	Signal Artifact/Noise	12/10/2023	12/10/2023
									A12	Connection Problem	12/10/2023	12/10/2023
							F1001	Absence of Treatment	A070909	Oversensing	12/10/2023	12/10/2023
									A0712	Pacing Problem	12/10/2023	12/10/2023
									A072201	High Impedance	12/10/2023	12/10/2023
									A090801	Signal Artifact/Noise	12/10/2023	12/10/2023
									A12	Connection Problem	12/10/2023	12/10/2023
							F1905	Device Revision or Replacement	A070909	Oversensing	12/10/2023	12/10/2023
									A0712	Pacing Problem	12/10/2023	12/10/2023
									A072201	High Impedance	12/10/2023	12/10/2023
									A090801	Signal Artifact/Noise	12/10/2023	12/10/2023
									A12	Connection Problem	12/10/2023	12/10/2023
					CT257	Implantable cardiac pacemakers	E011902	Presyncope	A070909	Oversensing	12/10/2023	12/10/2023
									A0712	Pacing Problem	12/10/2023	12/10/2023
									A072201	High Impedance	12/10/2023	12/10/2023
									A090801	Signal Artifact/Noise	12/10/2023	12/10/2023
									A12	Connection Problem	12/10/2023	12/10/2023
							E060101	Asystole	A070909	Oversensing	12/10/2023	12/10/2023
									A0712	Pacing Problem	12/10/2023	12/10/2023
									A072201	High Impedance	12/10/2023	12/10/2023
									A090801	Signal Artifact/Noise	12/10/2023	12/10/2023
									A12	Connection Problem	12/10/2023	12/10/2023
							F08	Hospitalization or Prolonged Hospitalization	A070909	Oversensing	12/10/2023	12/10/2023
									A0712	Pacing Problem	12/10/2023	12/10/2023
									A072201	High Impedance	12/10/2023	12/10/2023
									A090801	Signal Artifact/Noise	12/10/2023	12/10/2023
									A12	Connection Problem	12/10/2023	12/10/2023
							F1001	Absence of Treatment	A070909	Oversensing	12/10/2023	12/10/2023
									A0712	Pacing Problem	12/10/2023	12/10/2023
									A072201	High Impedance	12/10/2023	12/10/2023
									A090801	Signal Artifact/Noise	12/10/2023	12/10/2023
									A12	Connection Problem	12/10/2023	12/10/2023
							F1905	Device Revision or Replacement	A070909	Oversensing	12/10/2023	12/10/2023
									A0712	Pacing Problem	12/10/2023	12/10/2023
									A072201	High Impedance	12/10/2023	12/10/2023
									A090801	Signal Artifact/Noise	12/10/2023	12/10/2023
									A12	Connection Problem	12/10/2023	12/10/2023
					CT2866	Implantable cardiac electrical stimulators	E011902	Presyncope	A070909	Oversensing	12/10/2023	12/10/2023
									A0712	Pacing Problem	12/10/2023	12/10/2023
									A072201	High Impedance	12/10/2023	12/10/2023
									A090801	Signal Artifact/Noise	12/10/2023	12/10/2023
									A12	Connection Problem	12/10/2023	12/10/2023
							E060101	Asystole	A070909	Oversensing	12/10/2023	12/10/2023
									A0712	Pacing Problem	12/10/2023	12/10/2023
									A072201	High Impedance	12/10/2023	12/10/2023
									A090801	Signal Artifact/Noise	12/10/2023	12/10/2023
									A12	Connection Problem	12/10/2023	12/10/2023
							F08	Hospitalization or Prolonged Hospitalization	A070909	Oversensing	12/10/2023	12/10/2023
									A0712	Pacing Problem	12/10/2023	12/10/2023
									A072201	High Impedance	12/10/2023	12/10/2023
									A090801	Signal Artifact/Noise	12/10/2023	12/10/2023
									A12	Connection Problem	12/10/2023	12/10/2023
							F1001	Absence of Treatment	A070909	Oversensing	12/10/2023	12/10/2023
									A0712	Pacing Problem	12/10/2023	12/10/2023
									A072201	High Impedance	12/10/2023	12/10/2023
									A090801	Signal Artifact/Noise	12/10/2023	12/10/2023
									A12	Connection Problem	12/10/2023	12/10/2023
							F1905	Device Revision or Replacement	A070909	Oversensing	12/10/2023	12/10/2023
									A0712	Pacing Problem	12/10/2023	12/10/2023
									A072201	High Impedance	12/10/2023	12/10/2023
									A090801	Signal Artifact/Noise	12/10/2023	12/10/2023
									A12	Connection Problem	12/10/2023	12/10/2023
15246	2024	All other reportable incidents	47265	Dual-chamber implantable pacemaker, rate-responsive	CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	20/05/2024	20/05/2024
							F1901	Additional Surgery	A070504	Premature Discharge of Battery	20/05/2024	20/05/2024
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	20/05/2024	20/05/2024
							F1901	Additional Surgery	A070504	Premature Discharge of Battery	20/05/2024	20/05/2024
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	20/05/2024	20/05/2024
							F1901	Additional Surgery	A070504	Premature Discharge of Battery	20/05/2024	20/05/2024
15247	2024	All other reportable incidents			CT1684	Cardiac pacemakers and associated devices	E0112	Dizziness	A070101	Failure to Capture		
									A072201	High Impedance		
							E011903	Syncope/Fainting	A070101	Failure to Capture		
									A072201	High Impedance		
							F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A070101	Failure to Capture		
									A072201	High Impedance		
15248	2024	All other reportable incidents	17882	Physiologic-monitoring defibrillation system	CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A071204	No Pacing		
							F27	Problem Identified During Non-Clinical Procedure	A071204	No Pacing		
15249	2024	All other reportable incidents	47265	Dual-chamber implantable pacemaker, rate-responsive	CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A071102	Failure to Interrogate		
							F26	No Health Consequences or Impact	A071102	Failure to Interrogate		
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A071102	Failure to Interrogate		
							F26	No Health Consequences or Impact	A071102	Failure to Interrogate		
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A071102	Failure to Interrogate		
							F26	No Health Consequences or Impact	A071102	Failure to Interrogate		
15250	2024	All other reportable incidents	47265	Dual-chamber implantable pacemaker, rate-responsive	CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	23/05/2024	23/05/2024
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	23/05/2024	23/05/2024
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	23/05/2024	23/05/2024
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	23/05/2024	23/05/2024
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	23/05/2024	23/05/2024

							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	23/05/2024	23/05/2024
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	23/05/2024	23/05/2024
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	23/05/2024	23/05/2024
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	23/05/2024	23/05/2024
15251	2024	All other reportable incidents			CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
									A0908	Incorrect, Inadequate or Imprecise Result or Readings		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
									A0908	Incorrect, Inadequate or Imprecise Result or Readings		
15252	2024	All other reportable incidents	17882	Physiologic-monitoring defibrillation system	CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A0719	Unexpected Shutdown		
							F27	Problem Identified During Non-Clinical Procedure	A0719	Unexpected Shutdown		
15253	2024	All other reportable incidents	17882	Physiologic-monitoring defibrillation system	CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A0719	Unexpected Shutdown		
							F27	Problem Identified During Non-Clinical Procedure	A0719	Unexpected Shutdown		
15254	2024	All other reportable incidents	47205	Cardiac pulse generator programmer	CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A150105	Key or Button Unresponsive/not Working		
							F27	Problem Identified During Non-Clinical Procedure	A150105	Key or Button Unresponsive/not Working		
15255	2024	All other reportable incidents	35852	Single-chamber implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A0711	Interrogation Problem		
							F26	No Health Consequences or Impact	A0711	Interrogation Problem		
15256	2024		35852	Single-chamber implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F1901	Additional Surgery	A070504	Premature Discharge of Battery		
15257	2024	All other reportable incidents	37265	Dual-chamber implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A070909	Oversensing		
									A072201	High Impedance		
									A090801	Signal Artifact/Noise		
									A1107	Data Problem		
							F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A070909	Oversensing		
									A072201	High Impedance		
									A090801	Signal Artifact/Noise		
									A1107	Data Problem		
15258	2024	All other reportable incidents	47265	Dual-chamber implantable pacemaker, rate-responsive	CT1685	Cardiac pacemakers	E0112	Dizziness	A070101	Failure to Capture	21/05/2024	21/05/2024
									A0712	Pacing Problem	21/05/2024	21/05/2024
							E011903	Syncope/Fainting	A070101	Failure to Capture	21/05/2024	21/05/2024
									A0712	Pacing Problem	21/05/2024	21/05/2024
							E2007	Fall	A070101	Failure to Capture	21/05/2024	21/05/2024
									A0712	Pacing Problem	21/05/2024	21/05/2024
							F08	Hospitalization or Prolonged Hospitalization	A070101	Failure to Capture	21/05/2024	21/05/2024
									A0712	Pacing Problem	21/05/2024	21/05/2024
							F1203	Life Threatening Illness or Injury	A070101	Failure to Capture	21/05/2024	21/05/2024
									A0712	Pacing Problem	21/05/2024	21/05/2024
							F1905	Device Revision or Replacement	A070101	Failure to Capture	21/05/2024	21/05/2024
									A0712	Pacing Problem	21/05/2024	21/05/2024
							F2307	Reprogramming of Device	A070101	Failure to Capture	21/05/2024	21/05/2024
									A0712	Pacing Problem	21/05/2024	21/05/2024
					CT257	Implantable cardiac pacemakers	E0112	Dizziness	A070101	Failure to Capture	21/05/2024	21/05/2024
									A0712	Pacing Problem	21/05/2024	21/05/2024
							E011903	Syncope/Fainting	A070101	Failure to Capture	21/05/2024	21/05/2024
									A0712	Pacing Problem	21/05/2024	21/05/2024
							E2007	Fall	A070101	Failure to Capture	21/05/2024	21/05/2024
									A0712	Pacing Problem	21/05/2024	21/05/2024
							F08	Hospitalization or Prolonged Hospitalization	A070101	Failure to Capture	21/05/2024	21/05/2024
									A0712	Pacing Problem	21/05/2024	21/05/2024
							F1203	Life Threatening Illness or Injury	A070101	Failure to Capture	21/05/2024	21/05/2024
									A0712	Pacing Problem	21/05/2024	21/05/2024
							F1905	Device Revision or Replacement	A070101	Failure to Capture	21/05/2024	21/05/2024
									A0712	Pacing Problem	21/05/2024	21/05/2024
							F2307	Reprogramming of Device	A070101	Failure to Capture	21/05/2024	21/05/2024
									A0712	Pacing Problem	21/05/2024	21/05/2024
					CT2866	Implantable cardiac electrical stimulators	E0112	Dizziness	A070101	Failure to Capture	21/05/2024	21/05/2024
									A0712	Pacing Problem	21/05/2024	21/05/2024
							E011903	Syncope/Fainting	A070101	Failure to Capture	21/05/2024	21/05/2024
									A0712	Pacing Problem	21/05/2024	21/05/2024
							E2007	Fall	A070101	Failure to Capture	21/05/2024	21/05/2024
									A0712	Pacing Problem	21/05/2024	21/05/2024
							F08	Hospitalization or Prolonged Hospitalization	A070101	Failure to Capture	21/05/2024	21/05/2024
									A0712	Pacing Problem	21/05/2024	21/05/2024
							F1203	Life Threatening Illness or Injury	A070101	Failure to Capture	21/05/2024	21/05/2024
									A0712	Pacing Problem	21/05/2024	21/05/2024
							F1905	Device Revision or Replacement	A070101	Failure to Capture	21/05/2024	21/05/2024
									A0712	Pacing Problem	21/05/2024	21/05/2024
							F2307	Reprogramming of Device	A070101	Failure to Capture	21/05/2024	21/05/2024
									A0712	Pacing Problem	21/05/2024	21/05/2024
15259	2024	All other reportable incidents	47265	Dual-chamber implantable pacemaker, rate-responsive	CT141	Electrodes/Leads and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A072001	Electromagnetic Interference		
							F26	No Health Consequences or Impact	A072001	Electromagnetic Interference		
					CT145	Cardiovascular devices	E2403	No Clinical Signs, Symptoms or Conditions	A072001	Electromagnetic Interference		
							F26	No Health Consequences or Impact	A072001	Electromagnetic Interference		
					CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A072001	Electromagnetic Interference		
							F26	No Health Consequences or Impact	A072001	Electromagnetic Interference		
					CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A072001	Electromagnetic Interference		
							F26	No Health Consequences or Impact	A072001	Electromagnetic Interference		
					CT233	Surgical	E2403	No Clinical Signs, Symptoms or Conditions	A072001	Electromagnetic Interference		
							F26	No Health Consequences or Impact	A072001	Electromagnetic Interference		
					CT2406	Implantable	E2403	No Clinical Signs, Symptoms or Conditions	A072001	Electromagnetic Interference		
							F26	No Health Consequences or Impact	A072001	Electromagnetic Interference		
					CT2407	Active implantable	E2403	No Clinical Signs, Symptoms or Conditions	A072001	Electromagnetic Interference		
							F26	No Health Consequences or Impact	A072001	Electromagnetic Interference		
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A072001	Electromagnetic Interference		
							F26	No Health Consequences or Impact	A072001	Electromagnetic Interference		
					CT2727	Active implant devices/ancillaries	E2403	No Clinical Signs, Symptoms or Conditions	A072001	Electromagnetic Interference		
							F26	No Health Consequences or Impact	A072001	Electromagnetic Interference		

					CT2750	Implantable cardiac electrical stimulators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A072001	Electromagnetic Interference		
							F26	No Health Consequences or Impact	A072001	Electromagnetic Interference		
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A072001	Electromagnetic Interference		
							F26	No Health Consequences or Impact	A072001	Electromagnetic Interference		
					CT2985	Active	E2403	No Clinical Signs, Symptoms or Conditions	A072001	Electromagnetic Interference		
							F26	No Health Consequences or Impact	A072001	Electromagnetic Interference		
					CT3001	Implantable metal	E2403	No Clinical Signs, Symptoms or Conditions	A072001	Electromagnetic Interference		
							F26	No Health Consequences or Impact	A072001	Electromagnetic Interference		
					CT3002	Implantable synthetic polymer	E2403	No Clinical Signs, Symptoms or Conditions	A072001	Electromagnetic Interference		
							F26	No Health Consequences or Impact	A072001	Electromagnetic Interference		
					CT3006	Implantable inorganic material	E2403	No Clinical Signs, Symptoms or Conditions	A072001	Electromagnetic Interference		
							F26	No Health Consequences or Impact	A072001	Electromagnetic Interference		
					CT314	Electrically powered	E2403	No Clinical Signs, Symptoms or Conditions	A072001	Electromagnetic Interference		
							F26	No Health Consequences or Impact	A072001	Electromagnetic Interference		
					CT321	Long-term surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A072001	Electromagnetic Interference		
							F26	No Health Consequences or Impact	A072001	Electromagnetic Interference		
					CT449	Generators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A072001	Electromagnetic Interference		
							F26	No Health Consequences or Impact	A072001	Electromagnetic Interference		
					CT981	Single use	E2403	No Clinical Signs, Symptoms or Conditions	A072001	Electromagnetic Interference		
							F26	No Health Consequences or Impact	A072001	Electromagnetic Interference		
					CT983	Surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A072001	Electromagnetic Interference		
							F26	No Health Consequences or Impact	A072001	Electromagnetic Interference		
15260	2024	All other reportable incidents	35223	Endocardial/interventricular septal pacing lead	CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070909	Oversensing		
									A090801	Signal Artifact/Noise		
							F26	No Health Consequences or Impact	A070909	Oversensing		
									A090801	Signal Artifact/Noise		
15261	2024	Death	17882	Physiologic-monitoring defibrillation system	CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A07	Electrical /Electronic Property Problem		
							F02	Death	A07	Electrical /Electronic Property Problem		
15262	2024	All other reportable incidents	35852	Single-chamber implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E2401	Insufficient Information	A160602	Premature End-of-Life Indicator	15/05/2024	15/05/2024
							F1903	Device Explantation	A160602	Premature End-of-Life Indicator	15/05/2024	15/05/2024
15263	2024	All other reportable incidents	47265	Dual-chamber implantable pacemaker, rate-responsive	CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	01/05/2024	01/05/2024
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	01/05/2024	01/05/2024
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	01/05/2024	01/05/2024
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	01/05/2024	01/05/2024
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	01/05/2024	01/05/2024
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	01/05/2024	01/05/2024
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	01/05/2024	01/05/2024
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	01/05/2024	01/05/2024
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	01/05/2024	01/05/2024
15264	2024	All other reportable incidents	47263	Cardiac resynchronization therapy implantable pacemaker	CT1685	Cardiac pacemakers	E0601	Arrhythmia	A070504	Premature Discharge of Battery	17/05/2024	17/05/2024
									A071204	No Pacing	17/05/2024	17/05/2024
							F1901	Additional Surgery	A070504	Premature Discharge of Battery	17/05/2024	17/05/2024
									A071204	No Pacing	17/05/2024	17/05/2024
					CT257	Implantable cardiac pacemakers	E0601	Arrhythmia	A070504	Premature Discharge of Battery	17/05/2024	17/05/2024
									A071204	No Pacing	17/05/2024	17/05/2024
							F1901	Additional Surgery	A070504	Premature Discharge of Battery	17/05/2024	17/05/2024
									A071204	No Pacing	17/05/2024	17/05/2024
					CT2866	Implantable cardiac electrical stimulators	E0601	Arrhythmia	A070504	Premature Discharge of Battery	17/05/2024	17/05/2024
									A071204	No Pacing	17/05/2024	17/05/2024
							F1901	Additional Surgery	A070504	Premature Discharge of Battery	17/05/2024	17/05/2024
									A071204	No Pacing	17/05/2024	17/05/2024
15265	2024	All other reportable incidents	47205	Cardiac pulse generator programmer	CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A0902	Display or Visual Feedback Problem		
									A150105	Key or Button Unresponsive/not Working		
							F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A0902	Display or Visual Feedback Problem		
									A150105	Key or Button Unresponsive/not Working		
15266	2024	All other reportable incidents	35852	Single-chamber implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A0709	Device Sensing Problem		
							F12	Serious Injury/ Illness/ Impairment	A0709	Device Sensing Problem		
							F2307	Reprogramming of Device	A0709	Device Sensing Problem		
15267	2024	All other reportable incidents	47267	Single-chamber implantable pacemaker, rate-responsive	CT1685	Cardiac pacemakers	E0112	Dizziness	A04	Material Integrity Problem	12/07/2024	12/07/2024
									A070102	High Capture Threshold	12/07/2024	12/07/2024
									A070909	Oversensing	12/07/2024	12/07/2024
									A072201	High Impedance	12/07/2024	12/07/2024
									A072202	Low Impedance	12/07/2024	12/07/2024
									A090801	Signal Artifact/Noise	12/07/2024	12/07/2024
							F08	Hospitalization or Prolonged Hospitalization	A04	Material Integrity Problem	12/07/2024	12/07/2024
									A070102	High Capture Threshold	12/07/2024	12/07/2024
									A070909	Oversensing	12/07/2024	12/07/2024
									A072201	High Impedance	12/07/2024	12/07/2024
									A072202	Low Impedance	12/07/2024	12/07/2024
									A090801	Signal Artifact/Noise	12/07/2024	12/07/2024
							F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A04	Material Integrity Problem	12/07/2024	12/07/2024
									A070102	High Capture Threshold	12/07/2024	12/07/2024
									A070909	Oversensing	12/07/2024	12/07/2024
									A072201	High Impedance	12/07/2024	12/07/2024
									A072202	Low Impedance	12/07/2024	12/07/2024
									A090801	Signal Artifact/Noise	12/07/2024	12/07/2024
							F1905	Device Revision or Replacement	A04	Material Integrity Problem	12/07/2024	12/07/2024
									A070102	High Capture Threshold	12/07/2024	12/07/2024
									A070909	Oversensing	12/07/2024	12/07/2024
									A072201	High Impedance	12/07/2024	12/07/2024
									A072202	Low Impedance	12/07/2024	12/07/2024
									A090801	Signal Artifact/Noise	12/07/2024	12/07/2024
							F2203	Imaging Required	A04	Material Integrity Problem	12/07/2024	12/07/2024
									A070102	High Capture Threshold	12/07/2024	12/07/2024
									A070909	Oversensing	12/07/2024	12/07/2024
									A072201	High Impedance	12/07/2024	12/07/2024
									A072202	Low Impedance	12/07/2024	12/07/2024
									A090801	Signal Artifact/Noise	12/07/2024	12/07/2024

					CT257	Implantable cardiac pacemakers	E0112	Dizziness	A04	Material Integrity Problem	12/07/2024	12/07/2024
									A070102	High Capture Threshold	12/07/2024	12/07/2024
									A070909	Oversensing	12/07/2024	12/07/2024
									A072201	High Impedance	12/07/2024	12/07/2024
									A072202	Low Impedance	12/07/2024	12/07/2024
									A090801	Signal Artifact/Noise	12/07/2024	12/07/2024
							F08	Hospitalization or Prolonged Hospitalization	A04	Material Integrity Problem	12/07/2024	12/07/2024
									A070102	High Capture Threshold	12/07/2024	12/07/2024
									A070909	Oversensing	12/07/2024	12/07/2024
									A072201	High Impedance	12/07/2024	12/07/2024
									A072202	Low Impedance	12/07/2024	12/07/2024
									A090801	Signal Artifact/Noise	12/07/2024	12/07/2024
							F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A04	Material Integrity Problem	12/07/2024	12/07/2024
									A070102	High Capture Threshold	12/07/2024	12/07/2024
									A070909	Oversensing	12/07/2024	12/07/2024
									A072201	High Impedance	12/07/2024	12/07/2024
									A072202	Low Impedance	12/07/2024	12/07/2024
									A090801	Signal Artifact/Noise	12/07/2024	12/07/2024
							F1905	Device Revision or Replacement	A04	Material Integrity Problem	12/07/2024	12/07/2024
									A070102	High Capture Threshold	12/07/2024	12/07/2024
									A070909	Oversensing	12/07/2024	12/07/2024
									A072201	High Impedance	12/07/2024	12/07/2024
									A072202	Low Impedance	12/07/2024	12/07/2024
									A090801	Signal Artifact/Noise	12/07/2024	12/07/2024
							F2203	Imaging Required	A04	Material Integrity Problem	12/07/2024	12/07/2024
									A070102	High Capture Threshold	12/07/2024	12/07/2024
									A070909	Oversensing	12/07/2024	12/07/2024
									A072201	High Impedance	12/07/2024	12/07/2024
									A072202	Low Impedance	12/07/2024	12/07/2024
									A090801	Signal Artifact/Noise	12/07/2024	12/07/2024
					CT2866	Implantable cardiac electrical stimulators	E0112	Dizziness	A04	Material Integrity Problem	12/07/2024	12/07/2024
									A070102	High Capture Threshold	12/07/2024	12/07/2024
									A070909	Oversensing	12/07/2024	12/07/2024
									A072201	High Impedance	12/07/2024	12/07/2024
									A072202	Low Impedance	12/07/2024	12/07/2024
									A090801	Signal Artifact/Noise	12/07/2024	12/07/2024
							F08	Hospitalization or Prolonged Hospitalization	A04	Material Integrity Problem	12/07/2024	12/07/2024
									A070102	High Capture Threshold	12/07/2024	12/07/2024
									A070909	Oversensing	12/07/2024	12/07/2024
									A072201	High Impedance	12/07/2024	12/07/2024
									A072202	Low Impedance	12/07/2024	12/07/2024
									A090801	Signal Artifact/Noise	12/07/2024	12/07/2024
							F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A04	Material Integrity Problem	12/07/2024	12/07/2024
									A070102	High Capture Threshold	12/07/2024	12/07/2024
									A070909	Oversensing	12/07/2024	12/07/2024
									A072201	High Impedance	12/07/2024	12/07/2024
									A072202	Low Impedance	12/07/2024	12/07/2024
									A090801	Signal Artifact/Noise	12/07/2024	12/07/2024
							F1905	Device Revision or Replacement	A04	Material Integrity Problem	12/07/2024	12/07/2024
									A070102	High Capture Threshold	12/07/2024	12/07/2024
									A070909	Oversensing	12/07/2024	12/07/2024
									A072201	High Impedance	12/07/2024	12/07/2024
									A072202	Low Impedance	12/07/2024	12/07/2024
									A090801	Signal Artifact/Noise	12/07/2024	12/07/2024
							F2203	Imaging Required	A04	Material Integrity Problem	12/07/2024	12/07/2024
									A070102	High Capture Threshold	12/07/2024	12/07/2024
									A070909	Oversensing	12/07/2024	12/07/2024
									A072201	High Impedance	12/07/2024	12/07/2024
									A072202	Low Impedance	12/07/2024	12/07/2024
									A090801	Signal Artifact/Noise	12/07/2024	12/07/2024
15268	2024	All other reportable incidents	47265	Dual-chamber implantable pacemaker, rate-responsive	CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A0908	Incorrect, Inadequate or Imprecise Result or Readings	21/05/2024	21/05/2024
							F08	Hospitalization or Prolonged Hospitalization	A0908	Incorrect, Inadequate or Imprecise Result or Readings	21/05/2024	21/05/2024
							F1905	Device Revision or Replacement	A0908	Incorrect, Inadequate or Imprecise Result or Readings	21/05/2024	21/05/2024
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A0908	Incorrect, Inadequate or Imprecise Result or Readings	21/05/2024	21/05/2024
							F08	Hospitalization or Prolonged Hospitalization	A0908	Incorrect, Inadequate or Imprecise Result or Readings	21/05/2024	21/05/2024
							F1905	Device Revision or Replacement	A0908	Incorrect, Inadequate or Imprecise Result or Readings	21/05/2024	21/05/2024
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A0908	Incorrect, Inadequate or Imprecise Result or Readings	21/05/2024	21/05/2024
							F08	Hospitalization or Prolonged Hospitalization	A0908	Incorrect, Inadequate or Imprecise Result or Readings	21/05/2024	21/05/2024
							F1905	Device Revision or Replacement	A0908	Incorrect, Inadequate or Imprecise Result or Readings	21/05/2024	21/05/2024
15269	2024	All other reportable incidents	37265	Dual-chamber implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E060110	Ventricular Fibrillation	A071202	Inaccurate Synchronization		
							F1203	Life Threatening Illness or Injury	A071202	Inaccurate Synchronization		
							F15	Recognized Device or Procedural Complication	A071202	Inaccurate Synchronization		
15270	2024	All other reportable incidents	35852	Single-chamber implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E020201	Anxiety	A070504	Premature Discharge of Battery	22/02/2024	22/02/2024
									A0908	Incorrect, Inadequate or Imprecise Result or Readings	22/02/2024	22/02/2024
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	22/02/2024	22/02/2024
									A0908	Incorrect, Inadequate or Imprecise Result or Readings	22/02/2024	22/02/2024
							F11	Non-Serious Injury/ Illness/ Impairment	A070504	Premature Discharge of Battery	22/02/2024	22/02/2024
									A0908	Incorrect, Inadequate or Imprecise Result or Readings	22/02/2024	22/02/2024
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	22/02/2024	22/02/2024

									A0908	Incorrect, Inadequate or Imprecise Result or Readings	22/02/2024	22/02/2024
15271	2024	All other reportable incidents	47205	Cardiac pulse generator programmer	CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A0203	Defective Device		
							F27	Problem Identified During Non-Clinical Procedure	A0203	Defective Device		
15272	2024	All other reportable incidents	35852	Single-chamber implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	16/11/2023	16/11/2023
									A0908	Incorrect, Inadequate or Imprecise Result or Readings	16/11/2023	16/11/2023
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	16/11/2023	16/11/2023
									A0908	Incorrect, Inadequate or Imprecise Result or Readings	16/11/2023	16/11/2023
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	16/11/2023	16/11/2023
									A0908	Incorrect, Inadequate or Imprecise Result or Readings	16/11/2023	16/11/2023
15273	2024	All other reportable incidents	47263	Cardiac resynchronization therapy implantable pacemaker	CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A070909	Oversensing		
							F26	No Health Consequences or Impact	A070909	Oversensing		
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A070909	Oversensing		
							F26	No Health Consequences or Impact	A070909	Oversensing		
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A070909	Oversensing		
							F26	No Health Consequences or Impact	A070909	Oversensing		
15274	2024	All other reportable incidents	60789	Intracardiac pacemaker	CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A0701	Capturing Problem		
									A070102	High Capture Threshold		
							F2307	Reprogramming of Device	A0701	Capturing Problem		
									A070102	High Capture Threshold		
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A0701	Capturing Problem		
									A070102	High Capture Threshold		
							F2307	Reprogramming of Device	A0701	Capturing Problem		
									A070102	High Capture Threshold		
					CT2818	Catheterization/Cannulation/Fluid path devices	E2403	No Clinical Signs, Symptoms or Conditions	A0701	Capturing Problem		
									A070102	High Capture Threshold		
							F2307	Reprogramming of Device	A0701	Capturing Problem		
									A070102	High Capture Threshold		
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A0701	Capturing Problem		
									A070102	High Capture Threshold		
							F2307	Reprogramming of Device	A0701	Capturing Problem		
									A070102	High Capture Threshold		
15275	2024	All other reportable incidents	47205	Cardiac pulse generator programmer	CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A11	Computer Software Problem		
							F26	No Health Consequences or Impact	A11	Computer Software Problem		
15276	2024	All other reportable incidents	47265	Dual-chamber implantable pacemaker, rate-responsive	CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A05	Mechanical Problem	23/05/2024	23/05/2024
							F26	No Health Consequences or Impact	A05	Mechanical Problem	23/05/2024	23/05/2024
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A05	Mechanical Problem	23/05/2024	23/05/2024
							F26	No Health Consequences or Impact	A05	Mechanical Problem	23/05/2024	23/05/2024
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A05	Mechanical Problem	23/05/2024	23/05/2024
							F26	No Health Consequences or Impact	A05	Mechanical Problem	23/05/2024	23/05/2024
15277	2024	All other reportable incidents	47267	Single-chamber implantable pacemaker, rate-responsive	CT141	Electrodes/Leads and associated devices	E2311	Discomfort	A0723	Unexpected Mode Switch		
									A090407	Therapy Delivered to Incorrect Body Area		
							F23	Unexpected Medical Intervention	A0723	Unexpected Mode Switch		
									A090407	Therapy Delivered to Incorrect Body Area		
					CT145	Cardiovascular devices	E2311	Discomfort	A0723	Unexpected Mode Switch		
									A090407	Therapy Delivered to Incorrect Body Area		
							F23	Unexpected Medical Intervention	A0723	Unexpected Mode Switch		
									A090407	Therapy Delivered to Incorrect Body Area		
					CT1684	Cardiac pacemakers and associated devices	E2311	Discomfort	A0723	Unexpected Mode Switch		
									A090407	Therapy Delivered to Incorrect Body Area		
							F23	Unexpected Medical Intervention	A0723	Unexpected Mode Switch		
									A090407	Therapy Delivered to Incorrect Body Area		
					CT1685	Cardiac pacemakers	E2311	Discomfort	A0723	Unexpected Mode Switch		
									A090407	Therapy Delivered to Incorrect Body Area		
							F23	Unexpected Medical Intervention	A0723	Unexpected Mode Switch		
									A090407	Therapy Delivered to Incorrect Body Area		
					CT233	Surgical	E2311	Discomfort	A0723	Unexpected Mode Switch		
									A090407	Therapy Delivered to Incorrect Body Area		
							F23	Unexpected Medical Intervention	A0723	Unexpected Mode Switch		
									A090407	Therapy Delivered to Incorrect Body Area		
					CT2406	Implantable	E2311	Discomfort	A0723	Unexpected Mode Switch		
									A090407	Therapy Delivered to Incorrect Body Area		
							F23	Unexpected Medical Intervention	A0723	Unexpected Mode Switch		
									A090407	Therapy Delivered to Incorrect Body Area		
					CT2407	Active implantable	E2311	Discomfort	A0723	Unexpected Mode Switch		
									A090407	Therapy Delivered to Incorrect Body Area		
							F23	Unexpected Medical Intervention	A0723	Unexpected Mode Switch		
									A090407	Therapy Delivered to Incorrect Body Area		
					CT257	Implantable cardiac pacemakers	E2311	Discomfort	A0723	Unexpected Mode Switch		
									A090407	Therapy Delivered to Incorrect Body Area		
							F23	Unexpected Medical Intervention	A0723	Unexpected Mode Switch		
									A090407	Therapy Delivered to Incorrect Body Area		
					CT2727	Active implant devices/ancillaries	E2311	Discomfort	A0723	Unexpected Mode Switch		
									A090407	Therapy Delivered to Incorrect Body Area		
							F23	Unexpected Medical Intervention	A0723	Unexpected Mode Switch		
									A090407	Therapy Delivered to Incorrect Body Area		
					CT2750	Implantable cardiac electrical stimulators and associated devices	E2311	Discomfort	A0723	Unexpected Mode Switch		
									A090407	Therapy Delivered to Incorrect Body Area		
							F23	Unexpected Medical Intervention	A0723	Unexpected Mode Switch		
									A090407	Therapy Delivered to Incorrect Body Area		
					CT2866	Implantable cardiac electrical stimulators	E2311	Discomfort	A0723	Unexpected Mode Switch		
									A090407	Therapy Delivered to Incorrect Body Area		
							F23	Unexpected Medical Intervention	A0723	Unexpected Mode Switch		
									A090407	Therapy Delivered to Incorrect Body Area		
					CT2985	Active	E2311	Discomfort	A0723	Unexpected Mode Switch		
									A090407	Therapy Delivered to Incorrect Body Area		

							F23	Unexpected Medical Intervention	A0723	Unexpected Mode Switch		
									A090407	Therapy Delivered to Incorrect Body Area		
					CT3001	Implantable metal	E2311	Discomfort	A0723	Unexpected Mode Switch		
									A090407	Therapy Delivered to Incorrect Body Area		
							F23	Unexpected Medical Intervention	A0723	Unexpected Mode Switch		
									A090407	Therapy Delivered to Incorrect Body Area		
					CT3002	Implantable synthetic polymer	E2311	Discomfort	A0723	Unexpected Mode Switch		
									A090407	Therapy Delivered to Incorrect Body Area		
							F23	Unexpected Medical Intervention	A0723	Unexpected Mode Switch		
									A090407	Therapy Delivered to Incorrect Body Area		
					CT3006	Implantable inorganic material	E2311	Discomfort	A0723	Unexpected Mode Switch		
									A090407	Therapy Delivered to Incorrect Body Area		
							F23	Unexpected Medical Intervention	A0723	Unexpected Mode Switch		
									A090407	Therapy Delivered to Incorrect Body Area		
					CT314	Electrically powered	E2311	Discomfort	A0723	Unexpected Mode Switch		
									A090407	Therapy Delivered to Incorrect Body Area		
							F23	Unexpected Medical Intervention	A0723	Unexpected Mode Switch		
									A090407	Therapy Delivered to Incorrect Body Area		
					CT321	Long-term surgical invasive	E2311	Discomfort	A0723	Unexpected Mode Switch		
									A090407	Therapy Delivered to Incorrect Body Area		
							F23	Unexpected Medical Intervention	A0723	Unexpected Mode Switch		
									A090407	Therapy Delivered to Incorrect Body Area		
					CT449	Generators and associated devices	E2311	Discomfort	A0723	Unexpected Mode Switch		
									A090407	Therapy Delivered to Incorrect Body Area		
							F23	Unexpected Medical Intervention	A0723	Unexpected Mode Switch		
									A090407	Therapy Delivered to Incorrect Body Area		
					CT981	Single use	E2311	Discomfort	A0723	Unexpected Mode Switch		
									A090407	Therapy Delivered to Incorrect Body Area		
							F23	Unexpected Medical Intervention	A0723	Unexpected Mode Switch		
									A090407	Therapy Delivered to Incorrect Body Area		
					CT983	Surgical invasive	E2311	Discomfort	A0723	Unexpected Mode Switch		
									A090407	Therapy Delivered to Incorrect Body Area		
							F23	Unexpected Medical Intervention	A0723	Unexpected Mode Switch		
									A090407	Therapy Delivered to Incorrect Body Area		
15278	2024	All other reportable incidents	47265	Dual-chamber implantable pacemaker, rate-responsive	CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A071102	Failure to Interrogate	28/05/2024	28/05/2024
									A071204	No Pacing	28/05/2024	28/05/2024
							F1901	Additional Surgery	A071102	Failure to Interrogate	28/05/2024	28/05/2024
									A071204	No Pacing	28/05/2024	28/05/2024
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A071102	Failure to Interrogate	28/05/2024	28/05/2024
									A071204	No Pacing	28/05/2024	28/05/2024
							F1901	Additional Surgery	A071102	Failure to Interrogate	28/05/2024	28/05/2024
									A071204	No Pacing	28/05/2024	28/05/2024
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A071102	Failure to Interrogate	28/05/2024	28/05/2024
									A071204	No Pacing	28/05/2024	28/05/2024
							F1901	Additional Surgery	A071102	Failure to Interrogate	28/05/2024	28/05/2024
									A071204	No Pacing	28/05/2024	28/05/2024
15279	2024	All other reportable incidents	47265	Dual-chamber implantable pacemaker, rate-responsive	CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	14/05/2024	14/05/2024
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	14/05/2024	14/05/2024
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	14/05/2024	14/05/2024
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	14/05/2024	14/05/2024
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	14/05/2024	14/05/2024
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	14/05/2024	14/05/2024
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	14/05/2024	14/05/2024
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	14/05/2024	14/05/2024
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	14/05/2024	14/05/2024
15280	2024	All other reportable incidents	37265	Dual-chamber implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	22/05/2024	22/05/2024
									A071102	Failure to Interrogate	22/05/2024	22/05/2024
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	22/05/2024	22/05/2024
									A071102	Failure to Interrogate	22/05/2024	22/05/2024
							F12	Serious Injury/ Illness/ Impairment	A070504	Premature Discharge of Battery	22/05/2024	22/05/2024
									A071102	Failure to Interrogate	22/05/2024	22/05/2024
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	22/05/2024	22/05/2024
									A071102	Failure to Interrogate	22/05/2024	22/05/2024
15281	2024	All other reportable incidents	47270	Cardiac resynchronization therapy implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A070909	Oversensing		
									A090801	Signal Artifact/Noise		
							F26	No Health Consequences or Impact	A070909	Oversensing		
									A090801	Signal Artifact/Noise		
15282	2024	All other reportable incidents	47265	Dual-chamber implantable pacemaker, rate-responsive	CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	06/12/2023	06/12/2023
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	06/12/2023	06/12/2023
							F1903	Device Explantation	A070504	Premature Discharge of Battery	06/12/2023	06/12/2023
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	06/12/2023	06/12/2023
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	06/12/2023	06/12/2023
							F1903	Device Explantation	A070504	Premature Discharge of Battery	06/12/2023	06/12/2023
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	06/12/2023	06/12/2023
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	06/12/2023	06/12/2023
							F1903	Device Explantation	A070504	Premature Discharge of Battery	06/12/2023	06/12/2023
15283	2024	All other reportable incidents	37265	Dual-chamber implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A072001	Electromagnetic Interference		
									A090801	Signal Artifact/Noise		
							F26	No Health Consequences or Impact	A072001	Electromagnetic Interference		
									A090801	Signal Artifact/Noise		
15284	2024	All other reportable incidents	47265	Dual-chamber implantable pacemaker, rate-responsive	CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	17/05/2024	17/05/2024
									A071102	Failure to Interrogate	17/05/2024	17/05/2024
									A0723	Unexpected Mode Switch	17/05/2024	17/05/2024
							F1901	Additional Surgery	A070504	Premature Discharge of Battery	17/05/2024	17/05/2024
									A071102	Failure to Interrogate	17/05/2024	17/05/2024
									A0723	Unexpected Mode Switch	17/05/2024	17/05/2024
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	17/05/2024	17/05/2024
									A071102	Failure to Interrogate	17/05/2024	17/05/2024

									A0723	Unexpected Mode Switch	17/05/2024	17/05/2024
							F1901	Additional Surgery	A070504	Premature Discharge of Battery	17/05/2024	17/05/2024
									A071102	Failure to Interrogate	17/05/2024	17/05/2024
									A0723	Unexpected Mode Switch	17/05/2024	17/05/2024
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	17/05/2024	17/05/2024
									A071102	Failure to Interrogate	17/05/2024	17/05/2024
									A0723	Unexpected Mode Switch	17/05/2024	17/05/2024
							F1901	Additional Surgery	A070504	Premature Discharge of Battery	17/05/2024	17/05/2024
									A071102	Failure to Interrogate	17/05/2024	17/05/2024
									A0723	Unexpected Mode Switch	17/05/2024	17/05/2024
15285	2024	All other reportable incidents	47267	Single-chamber implantable pacemaker, rate-responsive	CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	14/01/2025	14/01/2025
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	14/01/2025	14/01/2025
							F1903	Device Explantation	A070504	Premature Discharge of Battery	14/01/2025	14/01/2025
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	14/01/2025	14/01/2025
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	14/01/2025	14/01/2025
							F1903	Device Explantation	A070504	Premature Discharge of Battery	14/01/2025	14/01/2025
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	14/01/2025	14/01/2025
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	14/01/2025	14/01/2025
							F1903	Device Explantation	A070504	Premature Discharge of Battery	14/01/2025	14/01/2025
15286	2024		47265	Dual-chamber implantable pacemaker, rate-responsive	CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	01/04/2024	30/04/2024
							F1901	Additional Surgery	A24	No Apparent Device Problem	01/04/2024	30/04/2024
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	01/04/2024	30/04/2024
							F1901	Additional Surgery	A24	No Apparent Device Problem	01/04/2024	30/04/2024
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	01/04/2024	30/04/2024
							F1901	Additional Surgery	A24	No Apparent Device Problem	01/04/2024	30/04/2024
15287	2024	All other reportable incidents	47265	Dual-chamber implantable pacemaker, rate-responsive	CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A070101	Failure to Capture		
									A070908	Failure to Sense		
									A0722	Impedance Problem		
							F1908	Prolonged Surgery	A070101	Failure to Capture		
									A070908	Failure to Sense		
									A0722	Impedance Problem		
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A070101	Failure to Capture		
									A070908	Failure to Sense		
									A0722	Impedance Problem		
							F1908	Prolonged Surgery	A070101	Failure to Capture		
									A070908	Failure to Sense		
									A0722	Impedance Problem		
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A070101	Failure to Capture		
									A070908	Failure to Sense		
									A0722	Impedance Problem		
							F1908	Prolonged Surgery	A070101	Failure to Capture		
									A070908	Failure to Sense		
									A0722	Impedance Problem		
15288	2024	All other reportable incidents	47270	Cardiac resynchronization therapy implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A09	Output Problem		
									A090811	Unable to Obtain Readings		
									A13	Communication or Transmission Problem		
							F26	No Health Consequences or Impact	A09	Output Problem		
									A090811	Unable to Obtain Readings		
									A13	Communication or Transmission Problem		
15289	2024	All other reportable incidents	35852	Single-chamber implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E060102	Atrial Fibrillation	A24	No Apparent Device Problem		
							F12	Serious Injury/ Illness/ Impairment	A24	No Apparent Device Problem		
							F2306	Resuscitation	A24	No Apparent Device Problem		
15290	2024	All other reportable incidents			CT1684	Cardiac pacemakers and associated devices	E0601	Arrhythmia	A070910	Undersensing		
									A0712	Pacing Problem		
									A071202	Inaccurate Synchronization		
							E0623	Unspecified Heart Problem	A070910	Undersensing		
									A0712	Pacing Problem		
									A071202	Inaccurate Synchronization		
							F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A070910	Undersensing		
									A0712	Pacing Problem		
									A071202	Inaccurate Synchronization		
15291	2024	All other reportable incidents	47265	Dual-chamber implantable pacemaker, rate-responsive	CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem	01/05/2024	01/05/2024
									A070504	Premature Discharge of Battery	01/05/2024	01/05/2024
									A072201	High Impedance	01/05/2024	01/05/2024
							F08	Hospitalization or Prolonged Hospitalization	A0705	Battery Problem	01/05/2024	01/05/2024
									A070504	Premature Discharge of Battery	01/05/2024	01/05/2024
									A072201	High Impedance	01/05/2024	01/05/2024
							F1905	Device Revision or Replacement	A0705	Battery Problem	01/05/2024	01/05/2024
									A070504	Premature Discharge of Battery	01/05/2024	01/05/2024
									A072201	High Impedance	01/05/2024	01/05/2024
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem	01/05/2024	01/05/2024
									A070504	Premature Discharge of Battery	01/05/2024	01/05/2024
									A072201	High Impedance	01/05/2024	01/05/2024
							F08	Hospitalization or Prolonged Hospitalization	A0705	Battery Problem	01/05/2024	01/05/2024
									A070504	Premature Discharge of Battery	01/05/2024	01/05/2024
									A072201	High Impedance	01/05/2024	01/05/2024
							F1905	Device Revision or Replacement	A0705	Battery Problem	01/05/2024	01/05/2024
									A070504	Premature Discharge of Battery	01/05/2024	01/05/2024
									A072201	High Impedance	01/05/2024	01/05/2024
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem	01/05/2024	01/05/2024
									A070504	Premature Discharge of Battery	01/05/2024	01/05/2024
									A072201	High Impedance	01/05/2024	01/05/2024
							F08	Hospitalization or Prolonged Hospitalization	A0705	Battery Problem	01/05/2024	01/05/2024
									A070504	Premature Discharge of Battery	01/05/2024	01/05/2024
									A072201	High Impedance	01/05/2024	01/05/2024
							F1905	Device Revision or Replacement	A0705	Battery Problem	01/05/2024	01/05/2024
									A070504	Premature Discharge of Battery	01/05/2024	01/05/2024
									A072201	High Impedance	01/05/2024	01/05/2024

15292	2024	All other reportable incidents	17882	Physiologic-monitoring defibrillation system	CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A090201	Device Displays Incorrect Message		
							F27	Problem Identified During Non-Clinical Procedure	A090201	Device Displays Incorrect Message		
15293	2024	All other reportable incidents	17882	Physiologic-monitoring defibrillation system	CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A07	Electrical /Electronic Property Problem		
							F27	Problem Identified During Non-Clinical Procedure	A07	Electrical /Electronic Property Problem		
15294	2024		47265	Dual-chamber implantable pacemaker, rate-responsive	CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	24/05/2024	24/05/2024
							F1901	Additional Surgery	A070504	Premature Discharge of Battery	24/05/2024	24/05/2024
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	24/05/2024	24/05/2024
							F1901	Additional Surgery	A070504	Premature Discharge of Battery	24/05/2024	24/05/2024
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	24/05/2024	24/05/2024
							F1901	Additional Surgery	A070504	Premature Discharge of Battery	24/05/2024	24/05/2024
15295	2024	All other reportable incidents	47267	Single-chamber implantable pacemaker, rate-responsive	CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A0908	Incorrect, Inadequate or Imprecise Result or Readings	22/05/2024	22/05/2024
							F08	Hospitalization or Prolonged Hospitalization	A0908	Incorrect, Inadequate or Imprecise Result or Readings	22/05/2024	22/05/2024
							F1905	Device Revision or Replacement	A0908	Incorrect, Inadequate or Imprecise Result or Readings	22/05/2024	22/05/2024
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A0908	Incorrect, Inadequate or Imprecise Result or Readings	22/05/2024	22/05/2024
							F08	Hospitalization or Prolonged Hospitalization	A0908	Incorrect, Inadequate or Imprecise Result or Readings	22/05/2024	22/05/2024
							F1905	Device Revision or Replacement	A0908	Incorrect, Inadequate or Imprecise Result or Readings	22/05/2024	22/05/2024
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A0908	Incorrect, Inadequate or Imprecise Result or Readings	22/05/2024	22/05/2024
							F08	Hospitalization or Prolonged Hospitalization	A0908	Incorrect, Inadequate or Imprecise Result or Readings	22/05/2024	22/05/2024
							F1905	Device Revision or Replacement	A0908	Incorrect, Inadequate or Imprecise Result or Readings	22/05/2024	22/05/2024
15296	2024	All other reportable incidents	17882	Physiologic-monitoring defibrillation system	CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A0719	Unexpected Shutdown		
							F27	Problem Identified During Non-Clinical Procedure	A0719	Unexpected Shutdown		
15297	2024	All other reportable incidents	63188	Left-ventricular wireless pacing system sensor/transmitter	CT145	Cardiovascular devices			A0708	Power Problem		
					CT1577	Wireless monitoring system transmitters			A0708	Power Problem		
					CT1684	Cardiac pacemakers and associated devices			A0708	Power Problem		
					CT2727	Active implant devices/ancillaries			A0708	Power Problem		
					CT2750	Implantable cardiac electrical stimulators and associated devices			A0708	Power Problem		
					CT2885	Physiological monitoring/measuring system transmitters/receivers/transceivers			A0708	Power Problem		
15298	2024	All other reportable incidents	36106	Implantable pacemaker analyser, intraoperative	CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A0902	Display or Visual Feedback Problem		
									A11	Computer Software Problem		
							F26	No Health Consequences or Impact	A0902	Display or Visual Feedback Problem		
									A11	Computer Software Problem		
15299	2024	All other reportable incidents	36106	Implantable pacemaker analyser, intraoperative	CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A0902	Display or Visual Feedback Problem		
									A11	Computer Software Problem		
							F27	Problem Identified During Non-Clinical Procedure	A0902	Display or Visual Feedback Problem		
									A11	Computer Software Problem		
15300	2024	All other reportable incidents	36106	Implantable pacemaker analyser, intraoperative	CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A0902	Display or Visual Feedback Problem		
									A13	Communication or Transmission Problem		
							F26	No Health Consequences or Impact	A0902	Display or Visual Feedback Problem		
									A13	Communication or Transmission Problem		
15301	2024	All other reportable incidents	36106	Implantable pacemaker analyser, intraoperative	CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A13	Communication or Transmission Problem		
							F26	No Health Consequences or Impact	A13	Communication or Transmission Problem		
15302	2024	All other reportable incidents			CT1684	Cardiac pacemakers and associated devices	E0112	Dizziness	A070504	Premature Discharge of Battery	13/05/2024	13/05/2024
									A070909	Oversensing	13/05/2024	13/05/2024
									A0712	Pacing Problem	13/05/2024	13/05/2024
									A0908	Incorrect, Inadequate or Imprecise Result or Readings	13/05/2024	13/05/2024
							E020204	Malaise	A070504	Premature Discharge of Battery	13/05/2024	13/05/2024
									A070909	Oversensing	13/05/2024	13/05/2024
									A0712	Pacing Problem	13/05/2024	13/05/2024
									A0908	Incorrect, Inadequate or Imprecise Result or Readings	13/05/2024	13/05/2024
							E060101	Asystole	A070504	Premature Discharge of Battery	13/05/2024	13/05/2024
									A070909	Oversensing	13/05/2024	13/05/2024
									A0712	Pacing Problem	13/05/2024	13/05/2024
									A0908	Incorrect, Inadequate or Imprecise Result or Readings	13/05/2024	13/05/2024
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	13/05/2024	13/05/2024
									A070909	Oversensing	13/05/2024	13/05/2024
									A0712	Pacing Problem	13/05/2024	13/05/2024
									A0908	Incorrect, Inadequate or Imprecise Result or Readings	13/05/2024	13/05/2024
							F1001	Absence of Treatment	A070504	Premature Discharge of Battery	13/05/2024	13/05/2024
									A070909	Oversensing	13/05/2024	13/05/2024
									A0712	Pacing Problem	13/05/2024	13/05/2024
									A0908	Incorrect, Inadequate or Imprecise Result or Readings	13/05/2024	13/05/2024
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	13/05/2024	13/05/2024
									A070909	Oversensing	13/05/2024	13/05/2024
									A0712	Pacing Problem	13/05/2024	13/05/2024
									A0908	Incorrect, Inadequate or Imprecise Result or Readings	13/05/2024	13/05/2024
							F2301	Additional Device Required	A070504	Premature Discharge of Battery	13/05/2024	13/05/2024
									A070909	Oversensing	13/05/2024	13/05/2024
									A0712	Pacing Problem	13/05/2024	13/05/2024
									A0908	Incorrect, Inadequate or Imprecise Result or Readings	13/05/2024	13/05/2024
					CT257	Implantable cardiac pacemakers	E0112	Dizziness	A070504	Premature Discharge of Battery	13/05/2024	13/05/2024
									A070909	Oversensing	13/05/2024	13/05/2024
									A0712	Pacing Problem	13/05/2024	13/05/2024
									A0908	Incorrect, Inadequate or Imprecise Result or Readings	13/05/2024	13/05/2024

							E020204	Malaise	A070504	Premature Discharge of Battery	13/05/2024	13/05/2024
									A070909	Oversensing	13/05/2024	13/05/2024
									A0712	Pacing Problem	13/05/2024	13/05/2024
									A0908	Incorrect, Inadequate or Imprecise Result or Readings	13/05/2024	13/05/2024
							E060101	Asystole	A070504	Premature Discharge of Battery	13/05/2024	13/05/2024
									A070909	Oversensing	13/05/2024	13/05/2024
									A0712	Pacing Problem	13/05/2024	13/05/2024
									A0908	Incorrect, Inadequate or Imprecise Result or Readings	13/05/2024	13/05/2024
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	13/05/2024	13/05/2024
									A070909	Oversensing	13/05/2024	13/05/2024
									A0712	Pacing Problem	13/05/2024	13/05/2024
									A0908	Incorrect, Inadequate or Imprecise Result or Readings	13/05/2024	13/05/2024
							F1001	Absence of Treatment	A070504	Premature Discharge of Battery	13/05/2024	13/05/2024
									A070909	Oversensing	13/05/2024	13/05/2024
									A0712	Pacing Problem	13/05/2024	13/05/2024
									A0908	Incorrect, Inadequate or Imprecise Result or Readings	13/05/2024	13/05/2024
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	13/05/2024	13/05/2024
									A070909	Oversensing	13/05/2024	13/05/2024
									A0712	Pacing Problem	13/05/2024	13/05/2024
									A0908	Incorrect, Inadequate or Imprecise Result or Readings	13/05/2024	13/05/2024
							F2301	Additional Device Required	A070504	Premature Discharge of Battery	13/05/2024	13/05/2024
									A070909	Oversensing	13/05/2024	13/05/2024
									A0712	Pacing Problem	13/05/2024	13/05/2024
									A0908	Incorrect, Inadequate or Imprecise Result or Readings	13/05/2024	13/05/2024
15303	2024	All other reportable incidents	35852	Single-chamber implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E2104	Electric Shock	A072201	High Impedance		
									A0908	Incorrect, Inadequate or Imprecise Result or Readings		
							F17	Sedation	A072201	High Impedance		
									A0908	Incorrect, Inadequate or Imprecise Result or Readings		
							F22	Unexpected Diagnostic Intervention	A072201	High Impedance		
									A0908	Incorrect, Inadequate or Imprecise Result or Readings		
15304	2024	All other reportable incidents	17882	Physiologic-monitoring defibrillation system	CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070803	Failure to Power Up		
							F27	Problem Identified During Non-Clinical Procedure	A070803	Failure to Power Up		
15305	2024	All other reportable incidents	17882	Physiologic-monitoring defibrillation system	CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A090201	Device Displays Incorrect Message		
							F26	No Health Consequences or Impact	A090201	Device Displays Incorrect Message		
15306	2024	All other reportable incidents	47205	Cardiac pulse generator programmer	CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A0902	Display or Visual Feedback Problem		
									A11	Computer Software Problem		
							F26	No Health Consequences or Impact	A0902	Display or Visual Feedback Problem		
									A11	Computer Software Problem		
15307	2024	All other reportable incidents	36106	Implantable pacemaker analyser, intraoperative	CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A13	Communication or Transmission Problem		
							F26	No Health Consequences or Impact	A13	Communication or Transmission Problem		
15308	2024		35822	External pacemaker, transcutaneous pacing	CT1685	Cardiac pacemakers	E2401	Insufficient Information	A0201	Product Quality Problem		
									A0202	Defective Component		
									A040104	Solder Joint Fracture		
							F2306	Resuscitation	A0201	Product Quality Problem		
									A0202	Defective Component		
									A040104	Solder Joint Fracture		
15309	2024		35822	External pacemaker, transcutaneous pacing	CT1685	Cardiac pacemakers	E060110	Ventricular Fibrillation	A0203	Defective Device		
							F23	Unexpected Medical Intervention	A0203	Defective Device		
15310	2024	All other reportable incidents	47267	Single-chamber implantable pacemaker, rate-responsive	CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A070909	Oversensing		
							F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A070909	Oversensing		
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A070909	Oversensing		
							F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A070909	Oversensing		
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A070909	Oversensing		
							F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A070909	Oversensing		
15311	2024	All other reportable incidents	47270	Cardiac resynchronization therapy implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A0203	Defective Device		
									A070909	Oversensing		
									A070910	Undersensing		
									A072202	Low Impedance		
									A090801	Signal Artifact/Noise		
									A1301	Failure to Read Input Signal		
							F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A0203	Defective Device		
									A070909	Oversensing		
									A070910	Undersensing		
									A072202	Low Impedance		
									A090801	Signal Artifact/Noise		
									A1301	Failure to Read Input Signal		
15312	2024	All other reportable incidents	47270	Cardiac resynchronization therapy implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A070909	Oversensing		
									A072201	High Impedance		
									A090801	Signal Artifact/Noise		
									A1301	Failure to Read Input Signal		
							F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A070909	Oversensing		
									A072201	High Impedance		
									A090801	Signal Artifact/Noise		
									A1301	Failure to Read Input Signal		
15313	2024	All other reportable incidents	35852	Single-chamber implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	02/07/2024	02/07/2024
									A071102	Failure to Interrogate	02/07/2024	02/07/2024
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	02/07/2024	02/07/2024
									A071102	Failure to Interrogate	02/07/2024	02/07/2024
							F12	Serious Injury/ Illness/ Impairment	A070504	Premature Discharge of Battery	02/07/2024	02/07/2024
									A071102	Failure to Interrogate	02/07/2024	02/07/2024
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	02/07/2024	02/07/2024

									A071102	Failure to Interrogate	02/07/2024	02/07/2024
15314	2024	All other reportable incidents	35852	Single-chamber implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E210401	Shock from Patient Lead(s)	A0709	Device Sensing Problem		
									A071302	Inappropriate/Inadequate Shock/Stimulation		
							F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A0709	Device Sensing Problem		
									A071302	Inappropriate/Inadequate Shock/Stimulation		
							F1203	Life Threatening Illness or Injury	A0709	Device Sensing Problem		
									A071302	Inappropriate/Inadequate Shock/Stimulation		
15315	2024	All other reportable incidents	17882	Physiologic-monitoring defibrillation system	CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A0719	Unexpected Shutdown		
							F27	Problem Identified During Non-Clinical Procedure	A0719	Unexpected Shutdown		
15316	2024	All other reportable incidents	47265	Dual-chamber implantable pacemaker, rate-responsive	CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	30/05/2024	30/05/2024
							F1901	Additional Surgery	A070504	Premature Discharge of Battery	30/05/2024	30/05/2024
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	30/05/2024	30/05/2024
							F1901	Additional Surgery	A070504	Premature Discharge of Battery	30/05/2024	30/05/2024
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	30/05/2024	30/05/2024
							F1901	Additional Surgery	A070504	Premature Discharge of Battery	30/05/2024	30/05/2024
15317	2024	All other reportable incidents			CT1684	Cardiac pacemakers and associated devices	E0601	Arrhythmia	A070910	Undersensing		
									A0712	Pacing Problem		
									A071202	Inaccurate Synchronization		
							E233001	Chest Pain	A070910	Undersensing		
									A0712	Pacing Problem		
									A071202	Inaccurate Synchronization		
							F08	Hospitalization or Prolonged Hospitalization	A070910	Undersensing		
									A0712	Pacing Problem		
									A071202	Inaccurate Synchronization		
							F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A070910	Undersensing		
									A0712	Pacing Problem		
									A071202	Inaccurate Synchronization		
15318	2024	All other reportable incidents	17882	Physiologic-monitoring defibrillation system	CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A0709	Device Sensing Problem		
							F27	Problem Identified During Non-Clinical Procedure	A0709	Device Sensing Problem		
15319	2024	All other reportable incidents	36106	Implantable pacemaker analyser, intraoperative	CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A09	Output Problem		
									A13	Communication or Transmission Problem		
							F26	No Health Consequences or Impact	A09	Output Problem		
									A13	Communication or Transmission Problem		
15320	2024	All other reportable incidents	47270	Cardiac resynchronization therapy implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E060109	Tachycardia	A0709	Device Sensing Problem		
									A071301	Failure to Deliver Shock/Stimulation		
							F1001	Absence of Treatment	A0709	Device Sensing Problem		
									A071301	Failure to Deliver Shock/Stimulation		
							F1203	Life Threatening Illness or Injury	A0709	Device Sensing Problem		
									A071301	Failure to Deliver Shock/Stimulation		
15321	2024	All other reportable incidents	37265	Dual-chamber implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E210401	Shock from Patient Lead(s)	A0709	Device Sensing Problem		
									A0712	Pacing Problem		
									A071302	Inappropriate/Inadequate Shock/Stimulation		
							F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A0709	Device Sensing Problem		
									A0712	Pacing Problem		
									A071302	Inappropriate/Inadequate Shock/Stimulation		
							F1203	Life Threatening Illness or Injury	A0709	Device Sensing Problem		
									A0712	Pacing Problem		
									A071302	Inappropriate/Inadequate Shock/Stimulation		
15322	2024	All other reportable incidents	35852	Single-chamber implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A072201	High Impedance		
							F26	No Health Consequences or Impact	A072201	High Impedance		
15323	2024	All other reportable incidents	37265	Dual-chamber implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A070909	Oversensing		
									A072201	High Impedance		
									A072202	Low Impedance		
									A090801	Signal Artifact/Noise		
							F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A070909	Oversensing		
									A072201	High Impedance		
									A072202	Low Impedance		
									A090801	Signal Artifact/Noise		
15324	2024	Unanticipated serious deterioration in state of health	47265	Dual-chamber implantable pacemaker, rate-responsive	CT1685	Cardiac pacemakers	E060101	Asystole	A05	Mechanical Problem	30/05/2024	30/05/2024
									A070101	Failure to Capture	30/05/2024	30/05/2024
							F08	Hospitalization or Prolonged Hospitalization	A05	Mechanical Problem	30/05/2024	30/05/2024
									A070101	Failure to Capture	30/05/2024	30/05/2024
							F1203	Life Threatening Illness or Injury	A05	Mechanical Problem	30/05/2024	30/05/2024
									A070101	Failure to Capture	30/05/2024	30/05/2024
							F1905	Device Revision or Replacement	A05	Mechanical Problem	30/05/2024	30/05/2024
									A070101	Failure to Capture	30/05/2024	30/05/2024
					CT257	Implantable cardiac pacemakers	E060101	Asystole	A05	Mechanical Problem	30/05/2024	30/05/2024
									A070101	Failure to Capture	30/05/2024	30/05/2024
							F08	Hospitalization or Prolonged Hospitalization	A05	Mechanical Problem	30/05/2024	30/05/2024
									A070101	Failure to Capture	30/05/2024	30/05/2024
							F1203	Life Threatening Illness or Injury	A05	Mechanical Problem	30/05/2024	30/05/2024
									A070101	Failure to Capture	30/05/2024	30/05/2024
							F1905	Device Revision or Replacement	A05	Mechanical Problem	30/05/2024	30/05/2024
									A070101	Failure to Capture	30/05/2024	30/05/2024
					CT2866	Implantable cardiac electrical stimulators	E060101	Asystole	A05	Mechanical Problem	30/05/2024	30/05/2024
									A070101	Failure to Capture	30/05/2024	30/05/2024
							F08	Hospitalization or Prolonged Hospitalization	A05	Mechanical Problem	30/05/2024	30/05/2024
									A070101	Failure to Capture	30/05/2024	30/05/2024
							F1203	Life Threatening Illness or Injury	A05	Mechanical Problem	30/05/2024	30/05/2024
									A070101	Failure to Capture	30/05/2024	30/05/2024
							F1905	Device Revision or Replacement	A05	Mechanical Problem	30/05/2024	30/05/2024
									A070101	Failure to Capture	30/05/2024	30/05/2024
15325	2024	All other reportable incidents	47265	Dual-chamber implantable pacemaker, rate-responsive	CT1685	Cardiac pacemakers	E0623	Unspecified Heart Problem	A071209	Extracardiac Stimulation	19/01/2024	19/01/2024
							E2103	Device Overstimulation of Tissue	A071209	Extracardiac Stimulation	19/01/2024	19/01/2024
							F08	Hospitalization or Prolonged Hospitalization	A071209	Extracardiac Stimulation	19/01/2024	19/01/2024
							F1905	Device Revision or Replacement	A071209	Extracardiac Stimulation	19/01/2024	19/01/2024
					CT257	Implantable cardiac pacemakers	E0623	Unspecified Heart Problem	A071209	Extracardiac Stimulation	19/01/2024	19/01/2024
							E2103	Device Overstimulation of Tissue	A071209	Extracardiac Stimulation	19/01/2024	19/01/2024

							F08	Hospitalization or Prolonged Hospitalization	A071209	Extracardiac Stimulation	19/01/2024	19/01/2024
							F1905	Device Revision or Replacement	A071209	Extracardiac Stimulation	19/01/2024	19/01/2024
					CT2866	Implantable cardiac electrical stimulators	E0623	Unspecified Heart Problem	A071209	Extracardiac Stimulation	19/01/2024	19/01/2024
							E2103	Device Overstimulation of Tissue	A071209	Extracardiac Stimulation	19/01/2024	19/01/2024
							F08	Hospitalization or Prolonged Hospitalization	A071209	Extracardiac Stimulation	19/01/2024	19/01/2024
							F1905	Device Revision or Replacement	A071209	Extracardiac Stimulation	19/01/2024	19/01/2024
15326	2024	All other reportable incidents	47270	Cardiac resynchronization therapy implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E210401	Shock from Patient Lead(s)	A0712	Pacing Problem		
									A071302	Inappropriate/Inadequate Shock/Stimulation		
							E2311	Discomfort	A0712	Pacing Problem		
									A071302	Inappropriate/Inadequate Shock/Stimulation		
							E2330	Pain	A0712	Pacing Problem		
									A071302	Inappropriate/Inadequate Shock/Stimulation		
							F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A0712	Pacing Problem		
									A071302	Inappropriate/Inadequate Shock/Stimulation		
							F23	Unexpected Medical Intervention	A0712	Pacing Problem		
									A071302	Inappropriate/Inadequate Shock/Stimulation		
15327	2024	All other reportable incidents	17882	Physiologic-monitoring defibrillation system	CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A07	Electrical /Electronic Property Problem		
							F27	Problem Identified During Non-Clinical Procedure	A07	Electrical /Electronic Property Problem		
15328	2024	All other reportable incidents	37265	Dual-chamber implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A0712	Pacing Problem	03/06/2024	03/06/2024
									A1206	Misconnection	03/06/2024	03/06/2024
							F12	Serious Injury/ Illness/ Impairment	A0712	Pacing Problem	03/06/2024	03/06/2024
									A1206	Misconnection	03/06/2024	03/06/2024
							F23	Unexpected Medical Intervention	A0712	Pacing Problem	03/06/2024	03/06/2024
									A1206	Misconnection	03/06/2024	03/06/2024
15329	2024		47265	Dual-chamber implantable pacemaker, rate-responsive	CT1685	Cardiac pacemakers	E011902	Presyncope	A070101	Failure to Capture		
									A071102	Failure to Interrogate		
									A071204	No Pacing		
							F1905	Device Revision or Replacement	A070101	Failure to Capture		
									A071102	Failure to Interrogate		
									A071204	No Pacing		
					CT257	Implantable cardiac pacemakers	E011902	Presyncope	A070101	Failure to Capture		
									A071102	Failure to Interrogate		
									A071204	No Pacing		
							F1905	Device Revision or Replacement	A070101	Failure to Capture		
									A071102	Failure to Interrogate		
									A071204	No Pacing		
					CT2866	Implantable cardiac electrical stimulators	E011902	Presyncope	A070101	Failure to Capture		
									A071102	Failure to Interrogate		
									A071204	No Pacing		
							F1905	Device Revision or Replacement	A070101	Failure to Capture		
									A071102	Failure to Interrogate		
									A071204	No Pacing		
15330	2024	All other reportable incidents	47263	Cardiac resynchronization therapy implantable pacemaker	CT1685	Cardiac pacemakers	E060101	Asystole	A070101	Failure to Capture		
									A070909	Oversensing		
									A0908	Incorrect, Inadequate or Imprecise Result or Readings		
							E0602	Cardiac Arrest	A070101	Failure to Capture		
									A070909	Oversensing		
									A0908	Incorrect, Inadequate or Imprecise Result or Readings		
							F08	Hospitalization or Prolonged Hospitalization	A070101	Failure to Capture		
									A070909	Oversensing		
									A0908	Incorrect, Inadequate or Imprecise Result or Readings		
							F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A070101	Failure to Capture		
									A070909	Oversensing		
									A0908	Incorrect, Inadequate or Imprecise Result or Readings		
							F1905	Device Revision or Replacement	A070101	Failure to Capture		
									A070909	Oversensing		
									A0908	Incorrect, Inadequate or Imprecise Result or Readings		
							F23	Unexpected Medical Intervention	A070101	Failure to Capture		
									A070909	Oversensing		
									A0908	Incorrect, Inadequate or Imprecise Result or Readings		
					CT257	Implantable cardiac pacemakers	E060101	Asystole	A070101	Failure to Capture		
									A070909	Oversensing		
									A0908	Incorrect, Inadequate or Imprecise Result or Readings		
							E0602	Cardiac Arrest	A070101	Failure to Capture		
									A070909	Oversensing		
									A0908	Incorrect, Inadequate or Imprecise Result or Readings		
							F08	Hospitalization or Prolonged Hospitalization	A070101	Failure to Capture		
									A070909	Oversensing		
									A0908	Incorrect, Inadequate or Imprecise Result or Readings		
							F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A070101	Failure to Capture		
									A070909	Oversensing		
									A0908	Incorrect, Inadequate or Imprecise Result or Readings		
							F1905	Device Revision or Replacement	A070101	Failure to Capture		
									A070909	Oversensing		
									A0908	Incorrect, Inadequate or Imprecise Result or Readings		
							F23	Unexpected Medical Intervention	A070101	Failure to Capture		
									A070909	Oversensing		
									A0908	Incorrect, Inadequate or Imprecise Result or Readings		
					CT2866	Implantable cardiac electrical stimulators	E060101	Asystole	A070101	Failure to Capture		

									A070909	Oversensing		
									A0908	Incorrect, Inadequate or Imprecise Result or Readings		
							E0602	Cardiac Arrest	A070101	Failure to Capture		
									A070909	Oversensing		
									A0908	Incorrect, Inadequate or Imprecise Result or Readings		
							F08	Hospitalization or Prolonged Hospitalization	A070101	Failure to Capture		
									A070909	Oversensing		
									A0908	Incorrect, Inadequate or Imprecise Result or Readings		
							F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A070101	Failure to Capture		
									A070909	Oversensing		
									A0908	Incorrect, Inadequate or Imprecise Result or Readings		
							F1905	Device Revision or Replacement	A070101	Failure to Capture		
									A070909	Oversensing		
									A0908	Incorrect, Inadequate or Imprecise Result or Readings		
							F23	Unexpected Medical Intervention	A070101	Failure to Capture		
									A070909	Oversensing		
									A0908	Incorrect, Inadequate or Imprecise Result or Readings		
15331	2024	All other reportable incidents	47270	Cardiac resynchronization therapy implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E060109	Tachycardia	A0709	Device Sensing Problem		
									A071301	Failure to Deliver Shock/Stimulation		
							F1001	Absence of Treatment	A0709	Device Sensing Problem		
									A071301	Failure to Deliver Shock/Stimulation		
							F1203	Life Threatening Illness or Injury	A0709	Device Sensing Problem		
									A071301	Failure to Deliver Shock/Stimulation		
15332	2024	All other reportable incidents	47940	Triple-chamber implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A0712	Pacing Problem		
							F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A0712	Pacing Problem		
15333	2024	All other reportable incidents	47265	Dual-chamber implantable pacemaker, rate-responsive	CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	15/03/2024	15/03/2024
							F1901	Additional Surgery	A24	No Apparent Device Problem	15/03/2024	15/03/2024
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	15/03/2024	15/03/2024
							F1901	Additional Surgery	A24	No Apparent Device Problem	15/03/2024	15/03/2024
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	15/03/2024	15/03/2024
							F1901	Additional Surgery	A24	No Apparent Device Problem	15/03/2024	15/03/2024
15334	2024	All other reportable incidents	17882	Physiologic-monitoring defibrillation system	CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem		
							F27	Problem Identified During Non-Clinical Procedure	A0705	Battery Problem		
15335	2024	All other reportable incidents	17882	Physiologic-monitoring defibrillation system	CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A0719	Unexpected Shutdown		
							F26	No Health Consequences or Impact	A0719	Unexpected Shutdown		
15336	2024	All other reportable incidents	37265	Dual-chamber implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A13	Communication or Transmission Problem		
							F12	Serious Injury/ Illness/ Impairment	A13	Communication or Transmission Problem		
							F2307	Reprogramming of Device	A13	Communication or Transmission Problem		
15337	2024	All other reportable incidents	47265	Dual-chamber implantable pacemaker, rate-responsive	CT1685	Cardiac pacemakers	E060101	Asystole	A070101	Failure to Capture	21/09/2023	21/09/2023
									A0908	Incorrect, Inadequate or Imprecise Result or Readings	21/09/2023	21/09/2023
							F08	Hospitalization or Prolonged Hospitalization	A070101	Failure to Capture	21/09/2023	21/09/2023
									A0908	Incorrect, Inadequate or Imprecise Result or Readings	21/09/2023	21/09/2023
							F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A070101	Failure to Capture	21/09/2023	21/09/2023
									A0908	Incorrect, Inadequate or Imprecise Result or Readings	21/09/2023	21/09/2023
							F1905	Device Revision or Replacement	A070101	Failure to Capture	21/09/2023	21/09/2023
									A0908	Incorrect, Inadequate or Imprecise Result or Readings	21/09/2023	21/09/2023
					CT257	Implantable cardiac pacemakers	E060101	Asystole	A070101	Failure to Capture	21/09/2023	21/09/2023
									A0908	Incorrect, Inadequate or Imprecise Result or Readings	21/09/2023	21/09/2023
							F08	Hospitalization or Prolonged Hospitalization	A070101	Failure to Capture	21/09/2023	21/09/2023
									A0908	Incorrect, Inadequate or Imprecise Result or Readings	21/09/2023	21/09/2023
							F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A070101	Failure to Capture	21/09/2023	21/09/2023
									A0908	Incorrect, Inadequate or Imprecise Result or Readings	21/09/2023	21/09/2023
							F1905	Device Revision or Replacement	A070101	Failure to Capture	21/09/2023	21/09/2023
									A0908	Incorrect, Inadequate or Imprecise Result or Readings	21/09/2023	21/09/2023
					CT2866	Implantable cardiac electrical stimulators	E060101	Asystole	A070101	Failure to Capture	21/09/2023	21/09/2023
									A0908	Incorrect, Inadequate or Imprecise Result or Readings	21/09/2023	21/09/2023
							F08	Hospitalization or Prolonged Hospitalization	A070101	Failure to Capture	21/09/2023	21/09/2023
									A0908	Incorrect, Inadequate or Imprecise Result or Readings	21/09/2023	21/09/2023
							F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A070101	Failure to Capture	21/09/2023	21/09/2023
									A0908	Incorrect, Inadequate or Imprecise Result or Readings	21/09/2023	21/09/2023
							F1905	Device Revision or Replacement	A070101	Failure to Capture	21/09/2023	21/09/2023
									A0908	Incorrect, Inadequate or Imprecise Result or Readings	21/09/2023	21/09/2023
15338	2024	All other reportable incidents	37265	Dual-chamber implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A070909	Oversensing		
									A0712	Pacing Problem		
							F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A070909	Oversensing		
									A0712	Pacing Problem		
							F15	Recognized Device or Procedural Complication	A070909	Oversensing		
									A0712	Pacing Problem		
15339	2024	All other reportable incidents	47267	Single-chamber implantable pacemaker, rate-responsive	CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem	05/06/2024	05/06/2024
									A070504	Premature Discharge of Battery	05/06/2024	05/06/2024
							F08	Hospitalization or Prolonged Hospitalization	A0705	Battery Problem	05/06/2024	05/06/2024
									A070504	Premature Discharge of Battery	05/06/2024	05/06/2024
							F1905	Device Revision or Replacement	A0705	Battery Problem	05/06/2024	05/06/2024
									A070504	Premature Discharge of Battery	05/06/2024	05/06/2024

					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem	05/06/2024	05/06/2024
									A070504	Premature Discharge of Battery	05/06/2024	05/06/2024
							F08	Hospitalization or Prolonged Hospitalization	A0705	Battery Problem	05/06/2024	05/06/2024
									A070504	Premature Discharge of Battery	05/06/2024	05/06/2024
							F1905	Device Revision or Replacement	A0705	Battery Problem	05/06/2024	05/06/2024
									A070504	Premature Discharge of Battery	05/06/2024	05/06/2024
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem	05/06/2024	05/06/2024
									A070504	Premature Discharge of Battery	05/06/2024	05/06/2024
							F08	Hospitalization or Prolonged Hospitalization	A0705	Battery Problem	05/06/2024	05/06/2024
									A070504	Premature Discharge of Battery	05/06/2024	05/06/2024
							F1905	Device Revision or Replacement	A0705	Battery Problem	05/06/2024	05/06/2024
									A070504	Premature Discharge of Battery	05/06/2024	05/06/2024
15340	2024	All other reportable incidents	47265	Dual-chamber implantable pacemaker, rate-responsive	CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem	12/02/2024	12/02/2024
									A070504	Premature Discharge of Battery	12/02/2024	12/02/2024
							F08	Hospitalization or Prolonged Hospitalization	A0705	Battery Problem	12/02/2024	12/02/2024
									A070504	Premature Discharge of Battery	12/02/2024	12/02/2024
							F1905	Device Revision or Replacement	A0705	Battery Problem	12/02/2024	12/02/2024
									A070504	Premature Discharge of Battery	12/02/2024	12/02/2024
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem	12/02/2024	12/02/2024
									A070504	Premature Discharge of Battery	12/02/2024	12/02/2024
							F08	Hospitalization or Prolonged Hospitalization	A0705	Battery Problem	12/02/2024	12/02/2024
									A070504	Premature Discharge of Battery	12/02/2024	12/02/2024
							F1905	Device Revision or Replacement	A0705	Battery Problem	12/02/2024	12/02/2024
									A070504	Premature Discharge of Battery	12/02/2024	12/02/2024
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem	12/02/2024	12/02/2024
									A070504	Premature Discharge of Battery	12/02/2024	12/02/2024
							F08	Hospitalization or Prolonged Hospitalization	A0705	Battery Problem	12/02/2024	12/02/2024
									A070504	Premature Discharge of Battery	12/02/2024	12/02/2024
							F1905	Device Revision or Replacement	A0705	Battery Problem	12/02/2024	12/02/2024
									A070504	Premature Discharge of Battery	12/02/2024	12/02/2024
15341	2024	All other reportable incidents	35852	Single-chamber implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E011903	Syncope/Fainting	A070909	Oversensing		
									A071201	Failure to Convert Rhythm		
									A071302	Inappropriate/Inadequate Shock/Stimulation		
							E210401	Shock from Patient Lead(s)	A070909	Oversensing		
									A071201	Failure to Convert Rhythm		
									A071302	Inappropriate/Inadequate Shock/Stimulation		
							F08	Hospitalization or Prolonged Hospitalization	A070909	Oversensing		
									A071201	Failure to Convert Rhythm		
									A071302	Inappropriate/Inadequate Shock/Stimulation		
							F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A070909	Oversensing		
									A071201	Failure to Convert Rhythm		
									A071302	Inappropriate/Inadequate Shock/Stimulation		
15342	2024	All other reportable incidents	47205	Cardiac pulse generator programmer	CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A13	Communication or Transmission Problem		
							F26	No Health Consequences or Impact	A13	Communication or Transmission Problem		
15343	2024	All other reportable incidents	47265	Dual-chamber implantable pacemaker, rate-responsive	CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A0908	Incorrect, Inadequate or Imprecise Result or Readings	23/05/2024	23/05/2024
							F08	Hospitalization or Prolonged Hospitalization	A0908	Incorrect, Inadequate or Imprecise Result or Readings	23/05/2024	23/05/2024
							F1905	Device Revision or Replacement	A0908	Incorrect, Inadequate or Imprecise Result or Readings	23/05/2024	23/05/2024
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A0908	Incorrect, Inadequate or Imprecise Result or Readings	23/05/2024	23/05/2024
							F08	Hospitalization or Prolonged Hospitalization	A0908	Incorrect, Inadequate or Imprecise Result or Readings	23/05/2024	23/05/2024
							F1905	Device Revision or Replacement	A0908	Incorrect, Inadequate or Imprecise Result or Readings	23/05/2024	23/05/2024
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A0908	Incorrect, Inadequate or Imprecise Result or Readings	23/05/2024	23/05/2024
							F08	Hospitalization or Prolonged Hospitalization	A0908	Incorrect, Inadequate or Imprecise Result or Readings	23/05/2024	23/05/2024
							F1905	Device Revision or Replacement	A0908	Incorrect, Inadequate or Imprecise Result or Readings	23/05/2024	23/05/2024
15344	2024	All other reportable incidents	47270	Cardiac resynchronization therapy implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A070909	Oversensing	23/05/2024	23/05/2024
							F08	Hospitalization or Prolonged Hospitalization	A070909	Oversensing	23/05/2024	23/05/2024
							F1905	Device Revision or Replacement	A070909	Oversensing	23/05/2024	23/05/2024
15345	2024	All other reportable incidents	47270	Cardiac resynchronization therapy implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A072201	High Impedance		
							F26	No Health Consequences or Impact	A072201	High Impedance		
15346	2024	All other reportable incidents	35852	Single-chamber implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A010402	Migration	10/03/2025	10/03/2025
									A0709	Device Sensing Problem	10/03/2025	10/03/2025
									A070909	Oversensing	10/03/2025	10/03/2025
									A072201	High Impedance	10/03/2025	10/03/2025
							F08	Hospitalization or Prolonged Hospitalization	A010402	Migration	10/03/2025	10/03/2025
									A0709	Device Sensing Problem	10/03/2025	10/03/2025
									A070909	Oversensing	10/03/2025	10/03/2025
									A072201	High Impedance	10/03/2025	10/03/2025
							F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A010402	Migration	10/03/2025	10/03/2025
									A0709	Device Sensing Problem	10/03/2025	10/03/2025
									A070909	Oversensing	10/03/2025	10/03/2025
									A072201	High Impedance	10/03/2025	10/03/2025
							F1905	Device Revision or Replacement	A010402	Migration	10/03/2025	10/03/2025
									A0709	Device Sensing Problem	10/03/2025	10/03/2025
									A070909	Oversensing	10/03/2025	10/03/2025
									A072201	High Impedance	10/03/2025	10/03/2025
							F2203	Imaging Required	A010402	Migration	10/03/2025	10/03/2025
									A0709	Device Sensing Problem	10/03/2025	10/03/2025
									A070909	Oversensing	10/03/2025	10/03/2025
									A072201	High Impedance	10/03/2025	10/03/2025
15347	2024	All other reportable incidents	47270	Cardiac resynchronization therapy implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A0203	Defective Device	10/03/2025	10/03/2025
									A0709	Device Sensing Problem	10/03/2025	10/03/2025
									A070909	Oversensing	10/03/2025	10/03/2025

									A072201	High Impedance	10/03/2025	10/03/2025
							F08	Hospitalization or Prolonged Hospitalization	A0203	Defective Device	10/03/2025	10/03/2025
									A0709	Device Sensing Problem	10/03/2025	10/03/2025
									A070909	Oversensing	10/03/2025	10/03/2025
									A072201	High Impedance	10/03/2025	10/03/2025
							F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A0203	Defective Device	10/03/2025	10/03/2025
									A0709	Device Sensing Problem	10/03/2025	10/03/2025
									A070909	Oversensing	10/03/2025	10/03/2025
									A072201	High Impedance	10/03/2025	10/03/2025
							F1905	Device Revision or Replacement	A0203	Defective Device	10/03/2025	10/03/2025
									A0709	Device Sensing Problem	10/03/2025	10/03/2025
									A070909	Oversensing	10/03/2025	10/03/2025
									A072201	High Impedance	10/03/2025	10/03/2025
							F2203	Imaging Required	A0203	Defective Device	10/03/2025	10/03/2025
									A0709	Device Sensing Problem	10/03/2025	10/03/2025
									A070909	Oversensing	10/03/2025	10/03/2025
									A072201	High Impedance	10/03/2025	10/03/2025
15348	2024	All other reportable incidents	35852	Single-chamber implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E210401	Shock from Patient Lead(s)	A0712	Pacing Problem		
									A071302	Inappropriate/Inadequate Shock/Stimulation		
							F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A0712	Pacing Problem		
									A071302	Inappropriate/Inadequate Shock/Stimulation		
15349	2024	All other reportable incidents	37265	Dual-chamber implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E060102	Atrial Fibrillation	A071202	Inaccurate Synchronization		
							F26	No Health Consequences or Impact	A071202	Inaccurate Synchronization		
15350	2024	All other reportable incidents	47265	Dual-chamber implantable pacemaker, rate-responsive	CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A0908	Incorrect, Inadequate or Imprecise Result or Readings	31/05/2024	31/05/2024
							F08	Hospitalization or Prolonged Hospitalization	A0908	Incorrect, Inadequate or Imprecise Result or Readings	31/05/2024	31/05/2024
							F1905	Device Revision or Replacement	A0908	Incorrect, Inadequate or Imprecise Result or Readings	31/05/2024	31/05/2024
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A0908	Incorrect, Inadequate or Imprecise Result or Readings	31/05/2024	31/05/2024
							F08	Hospitalization or Prolonged Hospitalization	A0908	Incorrect, Inadequate or Imprecise Result or Readings	31/05/2024	31/05/2024
							F1905	Device Revision or Replacement	A0908	Incorrect, Inadequate or Imprecise Result or Readings	31/05/2024	31/05/2024
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A0908	Incorrect, Inadequate or Imprecise Result or Readings	31/05/2024	31/05/2024
							F08	Hospitalization or Prolonged Hospitalization	A0908	Incorrect, Inadequate or Imprecise Result or Readings	31/05/2024	31/05/2024
							F1905	Device Revision or Replacement	A0908	Incorrect, Inadequate or Imprecise Result or Readings	31/05/2024	31/05/2024
15351	2024	All other reportable incidents	47270	Cardiac resynchronization therapy implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	17/06/2024	17/06/2024
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	17/06/2024	17/06/2024
							F12	Serious Injury/ Illness/ Impairment	A070504	Premature Discharge of Battery	17/06/2024	17/06/2024
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	17/06/2024	17/06/2024
15352	2024	All other reportable incidents	37265	Dual-chamber implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A070909	Oversensing		
									A072001	Electromagnetic Interference		
									A090801	Signal Artifact/Noise		
							F26	No Health Consequences or Impact	A070909	Oversensing		
									A072001	Electromagnetic Interference		
									A090801	Signal Artifact/Noise		
15353	2024	All other reportable incidents	35852	Single-chamber implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A090805	Incorrect Measurement	01/01/2024	31/12/2024
							F26	No Health Consequences or Impact	A090805	Incorrect Measurement	01/01/2024	31/12/2024
15354	2024	All other reportable incidents	17882	Physiologic-monitoring defibrillation system	CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A07	Electrical /Electronic Property Problem		
							F27	Problem Identified During Non-Clinical Procedure	A07	Electrical /Electronic Property Problem		
15355	2024	All other reportable incidents	47265	Dual-chamber implantable pacemaker, rate-responsive	CT1685	Cardiac pacemakers	E012207	Twitching	A0908	Incorrect, Inadequate or Imprecise Result or Readings	17/01/2024	17/01/2024
							F08	Hospitalization or Prolonged Hospitalization	A0908	Incorrect, Inadequate or Imprecise Result or Readings	17/01/2024	17/01/2024
							F1905	Device Revision or Replacement	A0908	Incorrect, Inadequate or Imprecise Result or Readings	17/01/2024	17/01/2024
					CT257	Implantable cardiac pacemakers	E012207	Twitching	A0908	Incorrect, Inadequate or Imprecise Result or Readings	17/01/2024	17/01/2024
							F08	Hospitalization or Prolonged Hospitalization	A0908	Incorrect, Inadequate or Imprecise Result or Readings	17/01/2024	17/01/2024
							F1905	Device Revision or Replacement	A0908	Incorrect, Inadequate or Imprecise Result or Readings	17/01/2024	17/01/2024
					CT2866	Implantable cardiac electrical stimulators	E012207	Twitching	A0908	Incorrect, Inadequate or Imprecise Result or Readings	17/01/2024	17/01/2024
							F08	Hospitalization or Prolonged Hospitalization	A0908	Incorrect, Inadequate or Imprecise Result or Readings	17/01/2024	17/01/2024
							F1905	Device Revision or Replacement	A0908	Incorrect, Inadequate or Imprecise Result or Readings	17/01/2024	17/01/2024
15356	2024	All other reportable incidents	47265	Dual-chamber implantable pacemaker, rate-responsive	CT1685	Cardiac pacemakers	E0601	Arrhythmia	A070504	Premature Discharge of Battery	06/06/2024	06/06/2024
									A071102	Failure to Interrogate	06/06/2024	06/06/2024
									A071204	No Pacing	06/06/2024	06/06/2024
							F12	Serious Injury/ Illness/ Impairment	A070504	Premature Discharge of Battery	06/06/2024	06/06/2024
									A071102	Failure to Interrogate	06/06/2024	06/06/2024
									A071204	No Pacing	06/06/2024	06/06/2024
					CT257	Implantable cardiac pacemakers	E0601	Arrhythmia	A070504	Premature Discharge of Battery	06/06/2024	06/06/2024
									A071102	Failure to Interrogate	06/06/2024	06/06/2024
									A071204	No Pacing	06/06/2024	06/06/2024
							F12	Serious Injury/ Illness/ Impairment	A070504	Premature Discharge of Battery	06/06/2024	06/06/2024
									A071102	Failure to Interrogate	06/06/2024	06/06/2024
									A071204	No Pacing	06/06/2024	06/06/2024
					CT2866	Implantable cardiac electrical stimulators	E0601	Arrhythmia	A070504	Premature Discharge of Battery	06/06/2024	06/06/2024
									A071102	Failure to Interrogate	06/06/2024	06/06/2024
									A071204	No Pacing	06/06/2024	06/06/2024
							F12	Serious Injury/ Illness/ Impairment	A070504	Premature Discharge of Battery	06/06/2024	06/06/2024
									A071102	Failure to Interrogate	06/06/2024	06/06/2024
									A071204	No Pacing	06/06/2024	06/06/2024
15357	2024		17882	Physiologic-monitoring defibrillation system	CT1684	Cardiac pacemakers and associated devices	E2401	Insufficient Information	A070908	Failure to Sense		

							F05	Delay to Treatment/ Therapy	A070908	Failure to Sense		
15358	2024	All other reportable incidents	47263	Cardiac resynchronization therapy implantable pacemaker	CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
15359	2024	All other reportable incidents	47265	Dual-chamber implantable pacemaker, rate-responsive	CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A0908	Incorrect, Inadequate or Imprecise Result or Readings	24/11/2023	24/11/2023
							F08	Hospitalization or Prolonged Hospitalization	A0908	Incorrect, Inadequate or Imprecise Result or Readings	24/11/2023	24/11/2023
							F1905	Device Revision or Replacement	A0908	Incorrect, Inadequate or Imprecise Result or Readings	24/11/2023	24/11/2023
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A0908	Incorrect, Inadequate or Imprecise Result or Readings	24/11/2023	24/11/2023
							F08	Hospitalization or Prolonged Hospitalization	A0908	Incorrect, Inadequate or Imprecise Result or Readings	24/11/2023	24/11/2023
							F1905	Device Revision or Replacement	A0908	Incorrect, Inadequate or Imprecise Result or Readings	24/11/2023	24/11/2023
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A0908	Incorrect, Inadequate or Imprecise Result or Readings	24/11/2023	24/11/2023
							F08	Hospitalization or Prolonged Hospitalization	A0908	Incorrect, Inadequate or Imprecise Result or Readings	24/11/2023	24/11/2023
							F1905	Device Revision or Replacement	A0908	Incorrect, Inadequate or Imprecise Result or Readings	24/11/2023	24/11/2023
15360	2024	All other reportable incidents	47265	Dual-chamber implantable pacemaker, rate-responsive	CT1685	Cardiac pacemakers	E0119	Loss of Consciousness	A07	Electrical /Electronic Property Problem		
							E2330	Pain	A07	Electrical /Electronic Property Problem		
							F1203	Life Threatening Illness or Injury	A07	Electrical /Electronic Property Problem		
					CT257	Implantable cardiac pacemakers	E0119	Loss of Consciousness	A07	Electrical /Electronic Property Problem		
							E2330	Pain	A07	Electrical /Electronic Property Problem		
							F1203	Life Threatening Illness or Injury	A07	Electrical /Electronic Property Problem		
					CT2866	Implantable cardiac electrical stimulators	E0119	Loss of Consciousness	A07	Electrical /Electronic Property Problem		
							E2330	Pain	A07	Electrical /Electronic Property Problem		
							F1203	Life Threatening Illness or Injury	A07	Electrical /Electronic Property Problem		
15361	2024	All other reportable incidents	17882	Physiologic-monitoring defibrillation system	CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070903	Failure to Analyze Signal		
							F26	No Health Consequences or Impact	A070903	Failure to Analyze Signal		
15362	2024	All other reportable incidents	35852	Single-chamber implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A071201	Failure to Convert Rhythm		
							F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A071201	Failure to Convert Rhythm		
15363	2024	All other reportable incidents	17882	Physiologic-monitoring defibrillation system	CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A12	Connection Problem		
							F27	Problem Identified During Non-Clinical Procedure	A12	Connection Problem		
15364	2024	All other reportable incidents	47270	Cardiac resynchronization therapy implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A072001	Electromagnetic Interference		
									A090801	Signal Artifact/Noise		
							F26	No Health Consequences or Impact	A072001	Electromagnetic Interference		
									A090801	Signal Artifact/Noise		
15365	2024	All other reportable incidents	47270	Cardiac resynchronization therapy implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
15366	2024	All other reportable incidents	47265	Dual-chamber implantable pacemaker, rate-responsive	CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A0712	Pacing Problem		
							F26	No Health Consequences or Impact	A0712	Pacing Problem		
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A0712	Pacing Problem		
							F26	No Health Consequences or Impact	A0712	Pacing Problem		
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A0712	Pacing Problem		
							F26	No Health Consequences or Impact	A0712	Pacing Problem		
15367	2024	All other reportable incidents	47265	Dual-chamber implantable pacemaker, rate-responsive	CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A0908	Incorrect, Inadequate or Imprecise Result or Readings	08/12/2023	08/12/2023
							F08	Hospitalization or Prolonged Hospitalization	A0908	Incorrect, Inadequate or Imprecise Result or Readings	08/12/2023	08/12/2023
							F1905	Device Revision or Replacement	A0908	Incorrect, Inadequate or Imprecise Result or Readings	08/12/2023	08/12/2023
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A0908	Incorrect, Inadequate or Imprecise Result or Readings	08/12/2023	08/12/2023
							F08	Hospitalization or Prolonged Hospitalization	A0908	Incorrect, Inadequate or Imprecise Result or Readings	08/12/2023	08/12/2023
							F1905	Device Revision or Replacement	A0908	Incorrect, Inadequate or Imprecise Result or Readings	08/12/2023	08/12/2023
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A0908	Incorrect, Inadequate or Imprecise Result or Readings	08/12/2023	08/12/2023
							F08	Hospitalization or Prolonged Hospitalization	A0908	Incorrect, Inadequate or Imprecise Result or Readings	08/12/2023	08/12/2023
							F1905	Device Revision or Replacement	A0908	Incorrect, Inadequate or Imprecise Result or Readings	08/12/2023	08/12/2023
15368	2024	All other reportable incidents	47263	Cardiac resynchronization therapy implantable pacemaker	CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A0908	Incorrect, Inadequate or Imprecise Result or Readings	07/06/2024	07/06/2024
							F08	Hospitalization or Prolonged Hospitalization	A0908	Incorrect, Inadequate or Imprecise Result or Readings	07/06/2024	07/06/2024
							F1905	Device Revision or Replacement	A0908	Incorrect, Inadequate or Imprecise Result or Readings	07/06/2024	07/06/2024
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A0908	Incorrect, Inadequate or Imprecise Result or Readings	07/06/2024	07/06/2024
							F08	Hospitalization or Prolonged Hospitalization	A0908	Incorrect, Inadequate or Imprecise Result or Readings	07/06/2024	07/06/2024
							F1905	Device Revision or Replacement	A0908	Incorrect, Inadequate or Imprecise Result or Readings	07/06/2024	07/06/2024
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A0908	Incorrect, Inadequate or Imprecise Result or Readings	07/06/2024	07/06/2024
							F08	Hospitalization or Prolonged Hospitalization	A0908	Incorrect, Inadequate or Imprecise Result or Readings	07/06/2024	07/06/2024
							F1905	Device Revision or Replacement	A0908	Incorrect, Inadequate or Imprecise Result or Readings	07/06/2024	07/06/2024
15369	2024	All other reportable incidents	17882	Physiologic-monitoring defibrillation system	CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A0706	Charging Problem		
							F27	Problem Identified During Non-Clinical Procedure	A0706	Charging Problem		
15370	2024	All other reportable incidents	47205	Cardiac pulse generator programmer	CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A150105	Key or Button Unresponsive/not Working		
							F26	No Health Consequences or Impact	A150105	Key or Button Unresponsive/not Working		
15371	2024	All other reportable incidents	47265	Dual-chamber implantable pacemaker, rate-responsive	CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A070910	Undersensing		
									A1301	Failure to Read Input Signal		

							F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A070910	Undersensing		
									A1301	Failure to Read Input Signal		
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A070910	Undersensing		
									A1301	Failure to Read Input Signal		
							F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A070910	Undersensing		
									A1301	Failure to Read Input Signal		
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A070910	Undersensing		
									A1301	Failure to Read Input Signal		
							F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A070910	Undersensing		
									A1301	Failure to Read Input Signal		
15372	2024	All other reportable incidents			CT1684	Cardiac pacemakers and associated devices	E2401	Insufficient Information	A0709	Device Sensing Problem		
							F1903	Device Explantation	A0709	Device Sensing Problem		
15373	2024	All other reportable incidents	37265	Dual-chamber implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E060109	Tachycardia	A0712	Pacing Problem		
							F1203	Life Threatening Illness or Injury	A0712	Pacing Problem		
15374	2024	All other reportable incidents	35852	Single-chamber implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	10/07/2024	10/07/2024
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	10/07/2024	10/07/2024
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	10/07/2024	10/07/2024
15375	2024	All other reportable incidents	17882	Physiologic-monitoring defibrillation system	CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A0719	Unexpected Shutdown		
							F27	Problem Identified During Non-Clinical Procedure	A0719	Unexpected Shutdown		
15376	2024	All other reportable incidents	17882	Physiologic-monitoring defibrillation system	CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A0719	Unexpected Shutdown		
							F26	No Health Consequences or Impact	A0719	Unexpected Shutdown		
15377	2024	All other reportable incidents	17882	Physiologic-monitoring defibrillation system	CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A071301	Failure to Deliver Shock/Stimulation		
									A071302	Inappropriate/Inadequate Shock/Stimulation		
							F27	Problem Identified During Non-Clinical Procedure	A071301	Failure to Deliver Shock/Stimulation		
									A071302	Inappropriate/Inadequate Shock/Stimulation		
15378	2024	All other reportable incidents	47265	Dual-chamber implantable pacemaker, rate-responsive	CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A160601	Premature Elective Replacement Indicator		
							F1905	Device Revision or Replacement	A160601	Premature Elective Replacement Indicator		
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A160601	Premature Elective Replacement Indicator		
							F1905	Device Revision or Replacement	A160601	Premature Elective Replacement Indicator		
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A160601	Premature Elective Replacement Indicator		
							F1905	Device Revision or Replacement	A160601	Premature Elective Replacement Indicator		
15379	2024	All other reportable incidents	36106	Implantable pacemaker analyser, intraoperative	CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A11	Computer Software Problem		
									A13	Communication or Transmission Problem		
							F26	No Health Consequences or Impact	A11	Computer Software Problem		
									A13	Communication or Transmission Problem		
15380	2024	All other reportable incidents	60789	Intracardiac pacemaker	CT1685	Cardiac pacemakers	E0605	Cardiac Tamponade	A24	No Apparent Device Problem		
							E0619	Pericardial Effusion	A24	No Apparent Device Problem		
							F08	Hospitalization or Prolonged Hospitalization	A24	No Apparent Device Problem		
							F12	Serious Injury/ Illness/ Impairment	A24	No Apparent Device Problem		
							F23	Unexpected Medical Intervention	A24	No Apparent Device Problem		
							F2303	Medication Required	A24	No Apparent Device Problem		
					CT257	Implantable cardiac pacemakers	E0605	Cardiac Tamponade	A24	No Apparent Device Problem		
							E0619	Pericardial Effusion	A24	No Apparent Device Problem		
							F08	Hospitalization or Prolonged Hospitalization	A24	No Apparent Device Problem		
							F12	Serious Injury/ Illness/ Impairment	A24	No Apparent Device Problem		
							F23	Unexpected Medical Intervention	A24	No Apparent Device Problem		
							F2303	Medication Required	A24	No Apparent Device Problem		
					CT2818	Catheterization/Cannulation/Fluid path devices	E0605	Cardiac Tamponade	A24	No Apparent Device Problem		
							E0619	Pericardial Effusion	A24	No Apparent Device Problem		
							F08	Hospitalization or Prolonged Hospitalization	A24	No Apparent Device Problem		
							F12	Serious Injury/ Illness/ Impairment	A24	No Apparent Device Problem		
							F23	Unexpected Medical Intervention	A24	No Apparent Device Problem		
							F2303	Medication Required	A24	No Apparent Device Problem		
					CT2866	Implantable cardiac electrical stimulators	E0605	Cardiac Tamponade	A24	No Apparent Device Problem		
							E0619	Pericardial Effusion	A24	No Apparent Device Problem		
							F08	Hospitalization or Prolonged Hospitalization	A24	No Apparent Device Problem		
							F12	Serious Injury/ Illness/ Impairment	A24	No Apparent Device Problem		
							F23	Unexpected Medical Intervention	A24	No Apparent Device Problem		
							F2303	Medication Required	A24	No Apparent Device Problem		
15381	2024	All other reportable incidents	17882	Physiologic-monitoring defibrillation system	CT1684	Cardiac pacemakers and associated devices	E2401	Insufficient Information	A26	Insufficient Device Problem Information		
							F26	No Health Consequences or Impact	A26	Insufficient Device Problem Information		
15382	2024	All other reportable incidents	47265	Dual-chamber implantable pacemaker, rate-responsive	CT1685	Cardiac pacemakers	E060101	Asystole	A070102	High Capture Threshold		
									A070909	Oversensing		
									A0712	Pacing Problem		
									A072201	High Impedance		
									A072202	Low Impedance		
									A090801	Signal Artifact/Noise		
							F1001	Absence of Treatment	A070102	High Capture Threshold		
									A070909	Oversensing		
									A0712	Pacing Problem		
									A072201	High Impedance		
									A072202	Low Impedance		
									A090801	Signal Artifact/Noise		
					CT257	Implantable cardiac pacemakers	E060101	Asystole	A070102	High Capture Threshold		
									A070909	Oversensing		
									A0712	Pacing Problem		
									A072201	High Impedance		
									A072202	Low Impedance		
									A090801	Signal Artifact/Noise		
							F1001	Absence of Treatment	A070102	High Capture Threshold		
									A070909	Oversensing		
									A0712	Pacing Problem		
									A072201	High Impedance		
									A072202	Low Impedance		
									A090801	Signal Artifact/Noise		
					CT2866	Implantable cardiac electrical stimulators	E060101	Asystole	A070102	High Capture Threshold		
									A070909	Oversensing		
									A0712	Pacing Problem		

									A072201	High Impedance		
									A072202	Low Impedance		
									A090801	Signal Artifact/Noise		
							F1001	Absence of Treatment	A070102	High Capture Threshold		
									A070909	Oversensing		
									A0712	Pacing Problem		
									A072201	High Impedance		
									A072202	Low Impedance		
									A090801	Signal Artifact/Noise		
15383	2024	All other reportable incidents	47270	Cardiac resynchronization therapy implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A070909	Oversensing		
									A071204	No Pacing		
							F26	No Health Consequences or Impact	A070909	Oversensing		
									A071204	No Pacing		
15384	2024	All other reportable incidents	47265	Dual-chamber implantable pacemaker, rate-responsive	CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A071102	Failure to Interrogate	14/06/2024	14/06/2024
									A071204	No Pacing	14/06/2024	14/06/2024
							F1901	Additional Surgery	A071102	Failure to Interrogate	14/06/2024	14/06/2024
									A071204	No Pacing	14/06/2024	14/06/2024
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A071102	Failure to Interrogate	14/06/2024	14/06/2024
									A071204	No Pacing	14/06/2024	14/06/2024
							F1901	Additional Surgery	A071102	Failure to Interrogate	14/06/2024	14/06/2024
									A071204	No Pacing	14/06/2024	14/06/2024
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A071102	Failure to Interrogate	14/06/2024	14/06/2024
									A071204	No Pacing	14/06/2024	14/06/2024
							F1901	Additional Surgery	A071102	Failure to Interrogate	14/06/2024	14/06/2024
									A071204	No Pacing	14/06/2024	14/06/2024
15385	2024	All other reportable incidents	37265	Dual-chamber implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A160601	Premature Elective Replacement Indicator		
							F1903	Device Explantation	A160601	Premature Elective Replacement Indicator		
15386	2024	All other reportable incidents	47265	Dual-chamber implantable pacemaker, rate-responsive	CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem	07/08/2024	07/08/2024
									A070504	Premature Discharge of Battery	07/08/2024	07/08/2024
							F08	Hospitalization or Prolonged Hospitalization	A0705	Battery Problem	07/08/2024	07/08/2024
									A070504	Premature Discharge of Battery	07/08/2024	07/08/2024
							F1905	Device Revision or Replacement	A0705	Battery Problem	07/08/2024	07/08/2024
									A070504	Premature Discharge of Battery	07/08/2024	07/08/2024
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem	07/08/2024	07/08/2024
									A070504	Premature Discharge of Battery	07/08/2024	07/08/2024
							F08	Hospitalization or Prolonged Hospitalization	A0705	Battery Problem	07/08/2024	07/08/2024
									A070504	Premature Discharge of Battery	07/08/2024	07/08/2024
							F1905	Device Revision or Replacement	A0705	Battery Problem	07/08/2024	07/08/2024
									A070504	Premature Discharge of Battery	07/08/2024	07/08/2024
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem	07/08/2024	07/08/2024
									A070504	Premature Discharge of Battery	07/08/2024	07/08/2024
							F08	Hospitalization or Prolonged Hospitalization	A0705	Battery Problem	07/08/2024	07/08/2024
									A070504	Premature Discharge of Battery	07/08/2024	07/08/2024
							F1905	Device Revision or Replacement	A0705	Battery Problem	07/08/2024	07/08/2024
									A070504	Premature Discharge of Battery	07/08/2024	07/08/2024
15387	2024	Death	37265	Dual-chamber implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E0601	Arrhythmia	A071201	Failure to Convert Rhythm		
									A071302	Inappropriate/Inadequate Shock/Stimulation		
							F02	Death	A071201	Failure to Convert Rhythm		
									A071302	Inappropriate/Inadequate Shock/Stimulation		
15388	2024	All other reportable incidents	47265	Dual-chamber implantable pacemaker, rate-responsive	CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A0908	Incorrect, Inadequate or Imprecise Result or Readings		
							F26	No Health Consequences or Impact	A0908	Incorrect, Inadequate or Imprecise Result or Readings		
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A0908	Incorrect, Inadequate or Imprecise Result or Readings		
							F26	No Health Consequences or Impact	A0908	Incorrect, Inadequate or Imprecise Result or Readings		
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A0908	Incorrect, Inadequate or Imprecise Result or Readings		
							F26	No Health Consequences or Impact	A0908	Incorrect, Inadequate or Imprecise Result or Readings		
15389	2024	All other reportable incidents	47263	Cardiac resynchronization therapy implantable pacemaker	CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A0908	Incorrect, Inadequate or Imprecise Result or Readings	18/06/2024	18/06/2024
							F08	Hospitalization or Prolonged Hospitalization	A0908	Incorrect, Inadequate or Imprecise Result or Readings	18/06/2024	18/06/2024
							F1905	Device Revision or Replacement	A0908	Incorrect, Inadequate or Imprecise Result or Readings	18/06/2024	18/06/2024
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A0908	Incorrect, Inadequate or Imprecise Result or Readings	18/06/2024	18/06/2024
							F08	Hospitalization or Prolonged Hospitalization	A0908	Incorrect, Inadequate or Imprecise Result or Readings	18/06/2024	18/06/2024
							F1905	Device Revision or Replacement	A0908	Incorrect, Inadequate or Imprecise Result or Readings	18/06/2024	18/06/2024
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A0908	Incorrect, Inadequate or Imprecise Result or Readings	18/06/2024	18/06/2024
							F08	Hospitalization or Prolonged Hospitalization	A0908	Incorrect, Inadequate or Imprecise Result or Readings	18/06/2024	18/06/2024
							F1905	Device Revision or Replacement	A0908	Incorrect, Inadequate or Imprecise Result or Readings	18/06/2024	18/06/2024
15390	2024	All other reportable incidents	37265	Dual-chamber implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A0712	Pacing Problem		
									A16	Protective Measures Problem		
							F26	No Health Consequences or Impact	A0712	Pacing Problem		
									A16	Protective Measures Problem		
15391	2024	All other reportable incidents	35852	Single-chamber implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E210401	Shock from Patient Lead(s)	A0709	Device Sensing Problem		
									A071302	Inappropriate/Inadequate Shock/Stimulation		
							F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A0709	Device Sensing Problem		
									A071302	Inappropriate/Inadequate Shock/Stimulation		
							F1203	Life Threatening Illness or Injury	A0709	Device Sensing Problem		
									A071302	Inappropriate/Inadequate Shock/Stimulation		
							F2307	Reprogramming of Device	A0709	Device Sensing Problem		
									A071302	Inappropriate/Inadequate Shock/Stimulation		

15392	2024	All other reportable incidents	47270	Cardiac resynchronization therapy implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A070909	Oversensing		
									A071204	No Pacing		
							F23	Unexpected Medical Intervention	A070909	Oversensing		
									A071204	No Pacing		
15393	2024	All other reportable incidents	37265	Dual-chamber implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A070909	Oversensing		
							F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A070909	Oversensing		
15394	2024	All other reportable incidents	47265	Dual-chamber implantable pacemaker, rate-responsive	CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A0908	Incorrect, Inadequate or Imprecise Result or Readings	14/03/2024	14/03/2024
							F08	Hospitalization or Prolonged Hospitalization	A0908	Incorrect, Inadequate or Imprecise Result or Readings	14/03/2024	14/03/2024
							F1905	Device Revision or Replacement	A0908	Incorrect, Inadequate or Imprecise Result or Readings	14/03/2024	14/03/2024
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A0908	Incorrect, Inadequate or Imprecise Result or Readings	14/03/2024	14/03/2024
							F08	Hospitalization or Prolonged Hospitalization	A0908	Incorrect, Inadequate or Imprecise Result or Readings	14/03/2024	14/03/2024
							F1905	Device Revision or Replacement	A0908	Incorrect, Inadequate or Imprecise Result or Readings	14/03/2024	14/03/2024
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A0908	Incorrect, Inadequate or Imprecise Result or Readings	14/03/2024	14/03/2024
							F08	Hospitalization or Prolonged Hospitalization	A0908	Incorrect, Inadequate or Imprecise Result or Readings	14/03/2024	14/03/2024
							F1905	Device Revision or Replacement	A0908	Incorrect, Inadequate or Imprecise Result or Readings	14/03/2024	14/03/2024
15395	2024	All other reportable incidents	47265	Dual-chamber implantable pacemaker, rate-responsive	CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A05	Mechanical Problem	03/04/2024	03/04/2024
							F08	Hospitalization or Prolonged Hospitalization	A05	Mechanical Problem	03/04/2024	03/04/2024
							F12	Serious Injury/ Illness/ Impairment	A05	Mechanical Problem	03/04/2024	03/04/2024
							F1905	Device Revision or Replacement	A05	Mechanical Problem	03/04/2024	03/04/2024
							F23	Unexpected Medical Intervention	A05	Mechanical Problem	03/04/2024	03/04/2024
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A05	Mechanical Problem	03/04/2024	03/04/2024
							F08	Hospitalization or Prolonged Hospitalization	A05	Mechanical Problem	03/04/2024	03/04/2024
							F12	Serious Injury/ Illness/ Impairment	A05	Mechanical Problem	03/04/2024	03/04/2024
							F1905	Device Revision or Replacement	A05	Mechanical Problem	03/04/2024	03/04/2024
							F23	Unexpected Medical Intervention	A05	Mechanical Problem	03/04/2024	03/04/2024
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A05	Mechanical Problem	03/04/2024	03/04/2024
							F08	Hospitalization or Prolonged Hospitalization	A05	Mechanical Problem	03/04/2024	03/04/2024
							F12	Serious Injury/ Illness/ Impairment	A05	Mechanical Problem	03/04/2024	03/04/2024
							F1905	Device Revision or Replacement	A05	Mechanical Problem	03/04/2024	03/04/2024
							F23	Unexpected Medical Intervention	A05	Mechanical Problem	03/04/2024	03/04/2024
15396	2024		47263	Cardiac resynchronization therapy implantable pacemaker	CT1685	Cardiac pacemakers	E012207	Twitching	A0203	Defective Device		
									A071205	Pacemaker Found in Back-Up Mode		
							F08	Hospitalization or Prolonged Hospitalization	A0203	Defective Device		
									A071205	Pacemaker Found in Back-Up Mode		
					CT257	Implantable cardiac pacemakers	E012207	Twitching	A0203	Defective Device		
									A071205	Pacemaker Found in Back-Up Mode		
							F08	Hospitalization or Prolonged Hospitalization	A0203	Defective Device		
									A071205	Pacemaker Found in Back-Up Mode		
					CT2866	Implantable cardiac electrical stimulators	E012207	Twitching	A0203	Defective Device		
									A071205	Pacemaker Found in Back-Up Mode		
							F08	Hospitalization or Prolonged Hospitalization	A0203	Defective Device		
									A071205	Pacemaker Found in Back-Up Mode		
15397	2024		37265	Dual-chamber implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A0202	Defective Component		
									A070602	Delayed Charge Time		
							F26	No Health Consequences or Impact	A0202	Defective Component		
									A070602	Delayed Charge Time		
15398	2024	All other reportable incidents	47265	Dual-chamber implantable pacemaker, rate-responsive	CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	06/06/2024	06/06/2024
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	06/06/2024	06/06/2024
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	06/06/2024	06/06/2024
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	06/06/2024	06/06/2024
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	06/06/2024	06/06/2024
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	06/06/2024	06/06/2024
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	06/06/2024	06/06/2024
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	06/06/2024	06/06/2024
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	06/06/2024	06/06/2024
15399	2024	All other reportable incidents	60789	Intracardiac pacemaker	CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
									A1605	Reset Problem		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
									A1605	Reset Problem		
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
									A1605	Reset Problem		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
									A1605	Reset Problem		
					CT2818	Catheterization/Cannulation/Fluid path devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
									A1605	Reset Problem		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
									A1605	Reset Problem		
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
									A1605	Reset Problem		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
									A1605	Reset Problem		
15400	2024		47265	Dual-chamber implantable pacemaker, rate-responsive	CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem		
							F26	No Health Consequences or Impact	A0705	Battery Problem		
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem		
							F26	No Health Consequences or Impact	A0705	Battery Problem		
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem		
							F26	No Health Consequences or Impact	A0705	Battery Problem		
15401	2024		47265	Dual-chamber implantable pacemaker, rate-responsive	CT1685	Cardiac pacemakers	E06	Heart	A07	Electrical /Electronic Property Problem		
									A0705	Battery Problem		
							E2403	No Clinical Signs, Symptoms or Conditions	A07	Electrical /Electronic Property Problem		
									A0705	Battery Problem		

							F26	No Health Consequences or Impact	A07	Electrical /Electronic Property Problem		
									A0705	Battery Problem		
					CT257	Implantable cardiac pacemakers	E06	Heart	A07	Electrical /Electronic Property Problem		
									A0705	Battery Problem		
							E2403	No Clinical Signs, Symptoms or Conditions	A07	Electrical /Electronic Property Problem		
									A0705	Battery Problem		
							F26	No Health Consequences or Impact	A07	Electrical /Electronic Property Problem		
									A0705	Battery Problem		
					CT2866	Implantable cardiac electrical stimulators	E06	Heart	A07	Electrical /Electronic Property Problem		
									A0705	Battery Problem		
							E2403	No Clinical Signs, Symptoms or Conditions	A07	Electrical /Electronic Property Problem		
									A0705	Battery Problem		
							F26	No Health Consequences or Impact	A07	Electrical /Electronic Property Problem		
									A0705	Battery Problem		
15402	2024		47270	Cardiac resynchronization therapy implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	A2301	Device Handling Problem	A0705	Battery Problem		
							E2401	Insufficient Information	A0705	Battery Problem		
							F26	No Health Consequences or Impact	A0705	Battery Problem		
15403	2024		47265	Dual-chamber implantable pacemaker, rate-responsive	CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem		
							F26	No Health Consequences or Impact	A0705	Battery Problem		
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem		
							F26	No Health Consequences or Impact	A0705	Battery Problem		
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem		
							F26	No Health Consequences or Impact	A0705	Battery Problem		
15404	2024		37265	Dual-chamber implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A0721	Circuit Failure		
							F26	No Health Consequences or Impact	A0721	Circuit Failure		
15405	2024	All other reportable incidents	47265	Dual-chamber implantable pacemaker, rate-responsive	CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A0908	Incorrect, Inadequate or Imprecise Result or Readings	11/06/2024	11/06/2024
							F08	Hospitalization or Prolonged Hospitalization	A0908	Incorrect, Inadequate or Imprecise Result or Readings	11/06/2024	11/06/2024
							F1905	Device Revision or Replacement	A0908	Incorrect, Inadequate or Imprecise Result or Readings	11/06/2024	11/06/2024
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A0908	Incorrect, Inadequate or Imprecise Result or Readings	11/06/2024	11/06/2024
							F08	Hospitalization or Prolonged Hospitalization	A0908	Incorrect, Inadequate or Imprecise Result or Readings	11/06/2024	11/06/2024
							F1905	Device Revision or Replacement	A0908	Incorrect, Inadequate or Imprecise Result or Readings	11/06/2024	11/06/2024
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A0908	Incorrect, Inadequate or Imprecise Result or Readings	11/06/2024	11/06/2024
							F08	Hospitalization or Prolonged Hospitalization	A0908	Incorrect, Inadequate or Imprecise Result or Readings	11/06/2024	11/06/2024
							F1905	Device Revision or Replacement	A0908	Incorrect, Inadequate or Imprecise Result or Readings	11/06/2024	11/06/2024
15406	2024	All other reportable incidents	60789	Intracardiac pacemaker	CT1685	Cardiac pacemakers	E0621	Valvular Insufficiency/ Regurgitation	A24	No Apparent Device Problem	04/06/2024	04/06/2024
							F1901	Additional Surgery	A24	No Apparent Device Problem	04/06/2024	04/06/2024
					CT257	Implantable cardiac pacemakers	E0621	Valvular Insufficiency/ Regurgitation	A24	No Apparent Device Problem	04/06/2024	04/06/2024
							F1901	Additional Surgery	A24	No Apparent Device Problem	04/06/2024	04/06/2024
					CT2818	Catheterization/Cannulation/Fluid path devices	E0621	Valvular Insufficiency/ Regurgitation	A24	No Apparent Device Problem	04/06/2024	04/06/2024
							F1901	Additional Surgery	A24	No Apparent Device Problem	04/06/2024	04/06/2024
					CT2866	Implantable cardiac electrical stimulators	E0621	Valvular Insufficiency/ Regurgitation	A24	No Apparent Device Problem	04/06/2024	04/06/2024
							F1901	Additional Surgery	A24	No Apparent Device Problem	04/06/2024	04/06/2024
15407	2024	All other reportable incidents	47265	Dual-chamber implantable pacemaker, rate-responsive	CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A05	Mechanical Problem		
									A1205	Loose or Intermittent Connection		
							F26	No Health Consequences or Impact	A05	Mechanical Problem		
									A1205	Loose or Intermittent Connection		
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A05	Mechanical Problem		
									A1205	Loose or Intermittent Connection		
							F26	No Health Consequences or Impact	A05	Mechanical Problem		
									A1205	Loose or Intermittent Connection		
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A05	Mechanical Problem		
									A1205	Loose or Intermittent Connection		
							F26	No Health Consequences or Impact	A05	Mechanical Problem		
									A1205	Loose or Intermittent Connection		
15408	2024	All other reportable incidents	37265	Dual-chamber implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
15409	2024	All other reportable incidents	47263	Cardiac resynchronization therapy implantable pacemaker	CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A13	Communication or Transmission Problem		
							F26	No Health Consequences or Impact	A13	Communication or Transmission Problem		
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A13	Communication or Transmission Problem		
							F26	No Health Consequences or Impact	A13	Communication or Transmission Problem		
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A13	Communication or Transmission Problem		
							F26	No Health Consequences or Impact	A13	Communication or Transmission Problem		
15410	2024	All other reportable incidents			CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A0712	Pacing Problem		
									A1107	Data Problem		
									A23	Use of Device Problem		
							F26	No Health Consequences or Impact	A0712	Pacing Problem		
									A1107	Data Problem		
									A23	Use of Device Problem		
15411	2024	All other reportable incidents	17882	Physiologic-monitoring defibrillation system	CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070803	Failure to Power Up		
							F27	Problem Identified During Non-Clinical Procedure	A070803	Failure to Power Up		
15412	2024	All other reportable incidents	60789	Intracardiac pacemaker	CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A070102	High Capture Threshold		
									A0712	Pacing Problem		
							F12	Serious Injury/ Illness/ Impairment	A070102	High Capture Threshold		
									A0712	Pacing Problem		
							F2307	Reprogramming of Device	A070102	High Capture Threshold		
									A0712	Pacing Problem		
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A070102	High Capture Threshold		
									A0712	Pacing Problem		
							F12	Serious Injury/ Illness/ Impairment	A070102	High Capture Threshold		
									A0712	Pacing Problem		
							F2307	Reprogramming of Device	A070102	High Capture Threshold		

									A0712	Pacing Problem		
					CT2818	Catheterization/Cannulation/Fluid path devices	E2403	No Clinical Signs, Symptoms or Conditions	A070102	High Capture Threshold		
									A0712	Pacing Problem		
							F12	Serious Injury/ Illness/ Impairment	A070102	High Capture Threshold		
									A0712	Pacing Problem		
							F2307	Reprogramming of Device	A070102	High Capture Threshold		
									A0712	Pacing Problem		
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A070102	High Capture Threshold		
									A0712	Pacing Problem		
							F12	Serious Injury/ Illness/ Impairment	A070102	High Capture Threshold		
									A0712	Pacing Problem		
							F2307	Reprogramming of Device	A070102	High Capture Threshold		
									A0712	Pacing Problem		
15413	2024	All other reportable incidents	47267	Single-chamber implantable pacemaker, rate-responsive	CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A071102	Failure to Interrogate		
									A0712	Pacing Problem		
							F2307	Reprogramming of Device	A071102	Failure to Interrogate		
									A0712	Pacing Problem		
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A071102	Failure to Interrogate		
									A0712	Pacing Problem		
							F2307	Reprogramming of Device	A071102	Failure to Interrogate		
									A0712	Pacing Problem		
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A071102	Failure to Interrogate		
									A0712	Pacing Problem		
							F2307	Reprogramming of Device	A071102	Failure to Interrogate		
									A0712	Pacing Problem		
15414	2024	All other reportable incidents	35852	Single-chamber implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A072001	Electromagnetic Interference		
									A090801	Signal Artifact/Noise		
							F12	Serious Injury/ Illness/ Impairment	A072001	Electromagnetic Interference		
									A090801	Signal Artifact/Noise		
							F2307	Reprogramming of Device	A072001	Electromagnetic Interference		
									A090801	Signal Artifact/Noise		
15415	2024	All other reportable incidents	17882	Physiologic-monitoring defibrillation system	CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A07	Electrical /Electronic Property Problem		
							F27	Problem Identified During Non-Clinical Procedure	A07	Electrical /Electronic Property Problem		
15416	2024	All other reportable incidents	17882	Physiologic-monitoring defibrillation system	CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070803	Failure to Power Up		
							F27	Problem Identified During Non-Clinical Procedure	A070803	Failure to Power Up		
15417	2024	All other reportable incidents	35852	Single-chamber implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A27	Appropriate Device Problem Term/Code not Available		
							F26	No Health Consequences or Impact	A27	Appropriate Device Problem Term/Code not Available		
15418	2024	All other reportable incidents	47270	Cardiac resynchronization therapy implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A070909	Oversensing		
							F1901	Additional Surgery	A070909	Oversensing		
15419	2024	All other reportable incidents	17882	Physiologic-monitoring defibrillation system	CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A150105	Key or Button Unresponsive/not Working		
							F27	Problem Identified During Non-Clinical Procedure	A150105	Key or Button Unresponsive/not Working		
15420	2024	All other reportable incidents			CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
15421	2024	All other reportable incidents	36106	Implantable pacemaker analyser, intraoperative	CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A1104	Computer Operating System Problem		
							F26	No Health Consequences or Impact	A1104	Computer Operating System Problem		
15422	2024	All other reportable incidents	47265	Dual-chamber implantable pacemaker, rate-responsive	CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
15423	2024	All other reportable incidents	47265	Dual-chamber implantable pacemaker, rate-responsive	CT1685	Cardiac pacemakers	E011903	Syncope/Fainting	A0908	Incorrect, Inadequate or Imprecise Result or Readings	30/10/2024	30/10/2024
							F08	Hospitalization or Prolonged Hospitalization	A0908	Incorrect, Inadequate or Imprecise Result or Readings	30/10/2024	30/10/2024
							F1905	Device Revision or Replacement	A0908	Incorrect, Inadequate or Imprecise Result or Readings	30/10/2024	30/10/2024
					CT257	Implantable cardiac pacemakers	E011903	Syncope/Fainting	A0908	Incorrect, Inadequate or Imprecise Result or Readings	30/10/2024	30/10/2024
							F08	Hospitalization or Prolonged Hospitalization	A0908	Incorrect, Inadequate or Imprecise Result or Readings	30/10/2024	30/10/2024
							F1905	Device Revision or Replacement	A0908	Incorrect, Inadequate or Imprecise Result or Readings	30/10/2024	30/10/2024
					CT2866	Implantable cardiac electrical stimulators	E011903	Syncope/Fainting	A0908	Incorrect, Inadequate or Imprecise Result or Readings	30/10/2024	30/10/2024
							F08	Hospitalization or Prolonged Hospitalization	A0908	Incorrect, Inadequate or Imprecise Result or Readings	30/10/2024	30/10/2024
							F1905	Device Revision or Replacement	A0908	Incorrect, Inadequate or Imprecise Result or Readings	30/10/2024	30/10/2024
15424	2024	All other reportable incidents	47265	Dual-chamber implantable pacemaker, rate-responsive	CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A0908	Incorrect, Inadequate or Imprecise Result or Readings	03/07/2024	03/07/2024
							F08	Hospitalization or Prolonged Hospitalization	A0908	Incorrect, Inadequate or Imprecise Result or Readings	03/07/2024	03/07/2024
							F1905	Device Revision or Replacement	A0908	Incorrect, Inadequate or Imprecise Result or Readings	03/07/2024	03/07/2024
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A0908	Incorrect, Inadequate or Imprecise Result or Readings	03/07/2024	03/07/2024
							F08	Hospitalization or Prolonged Hospitalization	A0908	Incorrect, Inadequate or Imprecise Result or Readings	03/07/2024	03/07/2024
							F1905	Device Revision or Replacement	A0908	Incorrect, Inadequate or Imprecise Result or Readings	03/07/2024	03/07/2024
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A0908	Incorrect, Inadequate or Imprecise Result or Readings	03/07/2024	03/07/2024
							F08	Hospitalization or Prolonged Hospitalization	A0908	Incorrect, Inadequate or Imprecise Result or Readings	03/07/2024	03/07/2024
							F1905	Device Revision or Replacement	A0908	Incorrect, Inadequate or Imprecise Result or Readings	03/07/2024	03/07/2024
15425	2024	All other reportable incidents	17882	Physiologic-monitoring defibrillation system	CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A0719	Unexpected Shutdown		

							F27	Problem Identified During Non-Clinical Procedure	A0719	Unexpected Shutdown		
15426	2024	All other reportable incidents	37265	Dual-chamber implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	27/08/2024	27/08/2024
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	27/08/2024	27/08/2024
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	27/08/2024	27/08/2024
15427	2024	All other reportable incidents	47265	Dual-chamber implantable pacemaker, rate-responsive	CT1685	Cardiac pacemakers	E0601	Arrhythmia	A070101	Failure to Capture		
									A070102	High Capture Threshold		
									A0712	Pacing Problem		
									A0909	Unexpected Therapeutic Results		
							F05	Delay to Treatment/ Therapy	A070101	Failure to Capture		
									A070102	High Capture Threshold		
									A0712	Pacing Problem		
									A0909	Unexpected Therapeutic Results		
							F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A070101	Failure to Capture		
									A070102	High Capture Threshold		
									A0712	Pacing Problem		
									A0909	Unexpected Therapeutic Results		
					CT257	Implantable cardiac pacemakers	E0601	Arrhythmia	A070101	Failure to Capture		
									A070102	High Capture Threshold		
									A0712	Pacing Problem		
									A0909	Unexpected Therapeutic Results		
							F05	Delay to Treatment/ Therapy	A070101	Failure to Capture		
									A070102	High Capture Threshold		
									A0712	Pacing Problem		
									A0909	Unexpected Therapeutic Results		
							F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A070101	Failure to Capture		
									A070102	High Capture Threshold		
									A0712	Pacing Problem		
									A0909	Unexpected Therapeutic Results		
					CT2866	Implantable cardiac electrical stimulators	E0601	Arrhythmia	A070101	Failure to Capture		
									A070102	High Capture Threshold		
									A0712	Pacing Problem		
									A0909	Unexpected Therapeutic Results		
							F05	Delay to Treatment/ Therapy	A070101	Failure to Capture		
									A070102	High Capture Threshold		
									A0712	Pacing Problem		
									A0909	Unexpected Therapeutic Results		
							F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A070101	Failure to Capture		
									A070102	High Capture Threshold		
									A0712	Pacing Problem		
									A0909	Unexpected Therapeutic Results		
15428	2024	All other reportable incidents	17882	Physiologic-monitoring defibrillation system	CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A0713	Defibrillation/Stimulation Problem		
							F27	Problem Identified During Non-Clinical Procedure	A0713	Defibrillation/Stimulation Problem		
15429	2024	All other reportable incidents	47265	Dual-chamber implantable pacemaker, rate-responsive	CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	02/04/2024	02/04/2024
							F1901	Additional Surgery	A24	No Apparent Device Problem	02/04/2024	02/04/2024
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	02/04/2024	02/04/2024
							F1901	Additional Surgery	A24	No Apparent Device Problem	02/04/2024	02/04/2024
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	02/04/2024	02/04/2024
							F1901	Additional Surgery	A24	No Apparent Device Problem	02/04/2024	02/04/2024
15430	2024	All other reportable incidents	47265	Dual-chamber implantable pacemaker, rate-responsive	CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A071102	Failure to Interrogate		
							F26	No Health Consequences or Impact	A071102	Failure to Interrogate		
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A071102	Failure to Interrogate		
							F26	No Health Consequences or Impact	A071102	Failure to Interrogate		
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A071102	Failure to Interrogate		
							F26	No Health Consequences or Impact	A071102	Failure to Interrogate		
15431	2024	All other reportable incidents	35852	Single-chamber implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E2104	Electric Shock	A070909	Oversensing	08/12/2023	08/12/2023
									A071302	Inappropriate/Inadequate Shock/Stimulation	08/12/2023	08/12/2023
									A090801	Signal Artifact/Noise	08/12/2023	08/12/2023
							F08	Hospitalization or Prolonged Hospitalization	A070909	Oversensing	08/12/2023	08/12/2023
									A071302	Inappropriate/Inadequate Shock/Stimulation	08/12/2023	08/12/2023
									A090801	Signal Artifact/Noise	08/12/2023	08/12/2023
							F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A070909	Oversensing	08/12/2023	08/12/2023
									A071302	Inappropriate/Inadequate Shock/Stimulation	08/12/2023	08/12/2023
									A090801	Signal Artifact/Noise	08/12/2023	08/12/2023
							F1903	Device Explantation	A070909	Oversensing	08/12/2023	08/12/2023
									A071302	Inappropriate/Inadequate Shock/Stimulation	08/12/2023	08/12/2023
									A090801	Signal Artifact/Noise	08/12/2023	08/12/2023
15432	2024	All other reportable incidents	47265	Dual-chamber implantable pacemaker, rate-responsive	CT1685	Cardiac pacemakers	E0601	Arrhythmia	A0723	Unexpected Mode Switch	06/03/2024	06/03/2024
							F1901	Additional Surgery	A0723	Unexpected Mode Switch	06/03/2024	06/03/2024
					CT257	Implantable cardiac pacemakers	E0601	Arrhythmia	A0723	Unexpected Mode Switch	06/03/2024	06/03/2024
							F1901	Additional Surgery	A0723	Unexpected Mode Switch	06/03/2024	06/03/2024
					CT2866	Implantable cardiac electrical stimulators	E0601	Arrhythmia	A0723	Unexpected Mode Switch	06/03/2024	06/03/2024
							F1901	Additional Surgery	A0723	Unexpected Mode Switch	06/03/2024	06/03/2024
15433	2024	All other reportable incidents	47265	Dual-chamber implantable pacemaker, rate-responsive	CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A0908	Incorrect, Inadequate or Imprecise Result or Readings	26/01/2024	26/01/2024
							F08	Hospitalization or Prolonged Hospitalization	A0908	Incorrect, Inadequate or Imprecise Result or Readings	26/01/2024	26/01/2024
							F1905	Device Revision or Replacement	A0908	Incorrect, Inadequate or Imprecise Result or Readings	26/01/2024	26/01/2024
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A0908	Incorrect, Inadequate or Imprecise Result or Readings	26/01/2024	26/01/2024
							F08	Hospitalization or Prolonged Hospitalization	A0908	Incorrect, Inadequate or Imprecise Result or Readings	26/01/2024	26/01/2024
							F1905	Device Revision or Replacement	A0908	Incorrect, Inadequate or Imprecise Result or Readings	26/01/2024	26/01/2024
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A0908	Incorrect, Inadequate or Imprecise Result or Readings	26/01/2024	26/01/2024
							F08	Hospitalization or Prolonged Hospitalization	A0908	Incorrect, Inadequate or Imprecise Result or Readings	26/01/2024	26/01/2024
							F1905	Device Revision or Replacement	A0908	Incorrect, Inadequate or Imprecise Result or Readings	26/01/2024	26/01/2024

15434	2024	All other reportable incidents	47267	Single-chamber implantable pacemaker, rate-responsive	CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	12/02/2024	12/02/2024
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	12/02/2024	12/02/2024
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	12/02/2024	12/02/2024
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	12/02/2024	12/02/2024
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	12/02/2024	12/02/2024
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	12/02/2024	12/02/2024
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	12/02/2024	12/02/2024
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	12/02/2024	12/02/2024
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	12/02/2024	12/02/2024
15435	2024	All other reportable incidents	47270	Cardiac resynchronization therapy implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A070909	Oversensing		
							F26	No Health Consequences or Impact	A070909	Oversensing		
15436	2024	All other reportable incidents	47267	Single-chamber implantable pacemaker, rate-responsive	CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem	15/11/2024	15/11/2024
									A070504	Premature Discharge of Battery	15/11/2024	15/11/2024
							F08	Hospitalization or Prolonged Hospitalization	A0705	Battery Problem	15/11/2024	15/11/2024
									A070504	Premature Discharge of Battery	15/11/2024	15/11/2024
							F1905	Device Revision or Replacement	A0705	Battery Problem	15/11/2024	15/11/2024
									A070504	Premature Discharge of Battery	15/11/2024	15/11/2024
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem	15/11/2024	15/11/2024
									A070504	Premature Discharge of Battery	15/11/2024	15/11/2024
							F08	Hospitalization or Prolonged Hospitalization	A0705	Battery Problem	15/11/2024	15/11/2024
									A070504	Premature Discharge of Battery	15/11/2024	15/11/2024
							F1905	Device Revision or Replacement	A0705	Battery Problem	15/11/2024	15/11/2024
									A070504	Premature Discharge of Battery	15/11/2024	15/11/2024
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem	15/11/2024	15/11/2024
									A070504	Premature Discharge of Battery	15/11/2024	15/11/2024
							F08	Hospitalization or Prolonged Hospitalization	A0705	Battery Problem	15/11/2024	15/11/2024
									A070504	Premature Discharge of Battery	15/11/2024	15/11/2024
							F1905	Device Revision or Replacement	A0705	Battery Problem	15/11/2024	15/11/2024
									A070504	Premature Discharge of Battery	15/11/2024	15/11/2024
15437	2024	All other reportable incidents	37265	Dual-chamber implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem	27/06/2024	27/06/2024
									A070504	Premature Discharge of Battery	27/06/2024	27/06/2024
									A0908	Incorrect, Inadequate or Imprecise Result or Readings	27/06/2024	27/06/2024
							F08	Hospitalization or Prolonged Hospitalization	A0705	Battery Problem	27/06/2024	27/06/2024
									A070504	Premature Discharge of Battery	27/06/2024	27/06/2024
									A0908	Incorrect, Inadequate or Imprecise Result or Readings	27/06/2024	27/06/2024
							F1905	Device Revision or Replacement	A0705	Battery Problem	27/06/2024	27/06/2024
									A070504	Premature Discharge of Battery	27/06/2024	27/06/2024
									A0908	Incorrect, Inadequate or Imprecise Result or Readings	27/06/2024	27/06/2024
15438	2024	All other reportable incidents	47270	Cardiac resynchronization therapy implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A072201	High Impedance		
							F26	No Health Consequences or Impact	A072201	High Impedance		
15439	2024		47265	Dual-chamber implantable pacemaker, rate-responsive	CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem		
							F1903	Device Explantation	A0705	Battery Problem		
							F26	No Health Consequences or Impact	A0705	Battery Problem		
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem		
							F1903	Device Explantation	A0705	Battery Problem		
							F26	No Health Consequences or Impact	A0705	Battery Problem		
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem		
							F1903	Device Explantation	A0705	Battery Problem		
							F26	No Health Consequences or Impact	A0705	Battery Problem		
15440	2024	All other reportable incidents	17882	Physiologic-monitoring defibrillation system	CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A0719	Unexpected Shutdown		
							F27	Problem Identified During Non-Clinical Procedure	A0719	Unexpected Shutdown		
15441	2024	All other reportable incidents	47270	Cardiac resynchronization therapy implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A070909	Oversensing		
							F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A070909	Oversensing		
15442	2024	All other reportable incidents	37265	Dual-chamber implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A072201	High Impedance		
							F08	Hospitalization or Prolonged Hospitalization	A072201	High Impedance		
							F19	Surgical Intervention	A072201	High Impedance		
15443	2024		35822	External pacemaker, transcutaneous pacing	CT1685	Cardiac pacemakers	E2402	Appropriate Clinical Signs, Symptoms and Conditions Term/Code not Available	A1603	Failure of Device to Self-Test		
							F2301	Additional Device Required	A1603	Failure of Device to Self-Test		
15444	2024	All other reportable incidents	47265	Dual-chamber implantable pacemaker, rate-responsive	CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A0908	Incorrect, Inadequate or Imprecise Result or Readings	09/09/2024	09/09/2024
							F08	Hospitalization or Prolonged Hospitalization	A0908	Incorrect, Inadequate or Imprecise Result or Readings	09/09/2024	09/09/2024
							F1903	Device Explantation	A0908	Incorrect, Inadequate or Imprecise Result or Readings	09/09/2024	09/09/2024
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A0908	Incorrect, Inadequate or Imprecise Result or Readings	09/09/2024	09/09/2024
							F08	Hospitalization or Prolonged Hospitalization	A0908	Incorrect, Inadequate or Imprecise Result or Readings	09/09/2024	09/09/2024
							F1903	Device Explantation	A0908	Incorrect, Inadequate or Imprecise Result or Readings	09/09/2024	09/09/2024
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A0908	Incorrect, Inadequate or Imprecise Result or Readings	09/09/2024	09/09/2024
							F08	Hospitalization or Prolonged Hospitalization	A0908	Incorrect, Inadequate or Imprecise Result or Readings	09/09/2024	09/09/2024
							F1903	Device Explantation	A0908	Incorrect, Inadequate or Imprecise Result or Readings	09/09/2024	09/09/2024
15445	2024	All other reportable incidents	17882	Physiologic-monitoring defibrillation system	CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070803	Failure to Power Up		
							F27	Problem Identified During Non-Clinical Procedure	A070803	Failure to Power Up		
15446	2024	All other reportable incidents	47270	Cardiac resynchronization therapy implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A070909	Oversensing		
									A071204	No Pacing		
							F1901	Additional Surgery	A070909	Oversensing		
									A071204	No Pacing		
15447	2024	All other reportable incidents	17882	Physiologic-monitoring defibrillation system	CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A07	Electrical /Electronic Property Problem		
							F27	Problem Identified During Non-Clinical Procedure	A07	Electrical /Electronic Property Problem		
15448	2024	All other reportable incidents	17882	Physiologic-monitoring defibrillation system	CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070803	Failure to Power Up		
							F27	Problem Identified During Non-Clinical Procedure	A070803	Failure to Power Up		
15449	2024	All other reportable incidents	17882	Physiologic-monitoring defibrillation system	CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A0719	Unexpected Shutdown		

							F27	Problem Identified During Non-Clinical Procedure	A0719	Unexpected Shutdown		
15450	2024	All other reportable incidents	47265	Dual-chamber implantable pacemaker, rate-responsive	CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A0712	Pacing Problem	23/06/2024	23/06/2024
							F08	Hospitalization or Prolonged Hospitalization	A0712	Pacing Problem	23/06/2024	23/06/2024
							F12	Serious Injury/ Illness/ Impairment	A0712	Pacing Problem	23/06/2024	23/06/2024
							F1905	Device Revision or Replacement	A0712	Pacing Problem	23/06/2024	23/06/2024
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A0712	Pacing Problem	23/06/2024	23/06/2024
							F08	Hospitalization or Prolonged Hospitalization	A0712	Pacing Problem	23/06/2024	23/06/2024
							F12	Serious Injury/ Illness/ Impairment	A0712	Pacing Problem	23/06/2024	23/06/2024
							F1905	Device Revision or Replacement	A0712	Pacing Problem	23/06/2024	23/06/2024
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A0712	Pacing Problem	23/06/2024	23/06/2024
							F08	Hospitalization or Prolonged Hospitalization	A0712	Pacing Problem	23/06/2024	23/06/2024
							F12	Serious Injury/ Illness/ Impairment	A0712	Pacing Problem	23/06/2024	23/06/2024
							F1905	Device Revision or Replacement	A0712	Pacing Problem	23/06/2024	23/06/2024
15451	2024	All other reportable incidents	17882	Physiologic-monitoring defibrillation system	CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070803	Failure to Power Up		
							F27	Problem Identified During Non-Clinical Procedure	A070803	Failure to Power Up		
15452	2024	All other reportable incidents	47267	Single-chamber implantable pacemaker, rate-responsive	CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem	14/06/2024	14/06/2024
									A070504	Premature Discharge of Battery	14/06/2024	14/06/2024
							F08	Hospitalization or Prolonged Hospitalization	A0705	Battery Problem	14/06/2024	14/06/2024
									A070504	Premature Discharge of Battery	14/06/2024	14/06/2024
							F1905	Device Revision or Replacement	A0705	Battery Problem	14/06/2024	14/06/2024
									A070504	Premature Discharge of Battery	14/06/2024	14/06/2024
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem	14/06/2024	14/06/2024
									A070504	Premature Discharge of Battery	14/06/2024	14/06/2024
							F08	Hospitalization or Prolonged Hospitalization	A0705	Battery Problem	14/06/2024	14/06/2024
									A070504	Premature Discharge of Battery	14/06/2024	14/06/2024
							F1905	Device Revision or Replacement	A0705	Battery Problem	14/06/2024	14/06/2024
									A070504	Premature Discharge of Battery	14/06/2024	14/06/2024
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem	14/06/2024	14/06/2024
									A070504	Premature Discharge of Battery	14/06/2024	14/06/2024
							F08	Hospitalization or Prolonged Hospitalization	A0705	Battery Problem	14/06/2024	14/06/2024
									A070504	Premature Discharge of Battery	14/06/2024	14/06/2024
							F1905	Device Revision or Replacement	A0705	Battery Problem	14/06/2024	14/06/2024
									A070504	Premature Discharge of Battery	14/06/2024	14/06/2024
15453	2024	All other reportable incidents	47270	Cardiac resynchronization therapy implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A0908	Incorrect, Inadequate or Imprecise Result or Readings		
							F26	No Health Consequences or Impact	A0908	Incorrect, Inadequate or Imprecise Result or Readings		
15454	2024	All other reportable incidents	35852	Single-chamber implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E210401	Shock from Patient Lead(s)	A070909	Oversensing		
									A071302	Inappropriate/Inadequate Shock/Stimulation		
							F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A070909	Oversensing		
									A071302	Inappropriate/Inadequate Shock/Stimulation		
							F2203	Imaging Required	A070909	Oversensing		
									A071302	Inappropriate/Inadequate Shock/Stimulation		
15455	2024	All other reportable incidents	37265	Dual-chamber implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E210401	Shock from Patient Lead(s)	A0712	Pacing Problem		
									A071302	Inappropriate/Inadequate Shock/Stimulation		
							F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A0712	Pacing Problem		
									A071302	Inappropriate/Inadequate Shock/Stimulation		
15456	2024	All other reportable incidents	35852	Single-chamber implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E2104	Electric Shock	A071302	Inappropriate/Inadequate Shock/Stimulation	28/11/2023	28/11/2023
									A0908	Incorrect, Inadequate or Imprecise Result or Readings	28/11/2023	28/11/2023
							F08	Hospitalization or Prolonged Hospitalization	A071302	Inappropriate/Inadequate Shock/Stimulation	28/11/2023	28/11/2023
									A0908	Incorrect, Inadequate or Imprecise Result or Readings	28/11/2023	28/11/2023
							F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A071302	Inappropriate/Inadequate Shock/Stimulation	28/11/2023	28/11/2023
									A0908	Incorrect, Inadequate or Imprecise Result or Readings	28/11/2023	28/11/2023
							F1905	Device Revision or Replacement	A071302	Inappropriate/Inadequate Shock/Stimulation	28/11/2023	28/11/2023
									A0908	Incorrect, Inadequate or Imprecise Result or Readings	28/11/2023	28/11/2023
							F23	Unexpected Medical Intervention	A071302	Inappropriate/Inadequate Shock/Stimulation	28/11/2023	28/11/2023
									A0908	Incorrect, Inadequate or Imprecise Result or Readings	28/11/2023	28/11/2023
15457	2024		37265	Dual-chamber implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A070909	Oversensing	06/02/2024	06/02/2024
							F1901	Additional Surgery	A070909	Oversensing	06/02/2024	06/02/2024
15458	2024	All other reportable incidents	47265	Dual-chamber implantable pacemaker, rate-responsive	CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A070909	Oversensing	15/05/2024	15/05/2024
									A0711	Interrogation Problem	15/05/2024	15/05/2024
									A072202	Low Impedance	15/05/2024	15/05/2024
									A090805	Incorrect Measurement	15/05/2024	15/05/2024
							F1901	Additional Surgery	A070909	Oversensing	15/05/2024	15/05/2024
									A0711	Interrogation Problem	15/05/2024	15/05/2024
									A072202	Low Impedance	15/05/2024	15/05/2024
									A090805	Incorrect Measurement	15/05/2024	15/05/2024
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A070909	Oversensing	15/05/2024	15/05/2024
									A0711	Interrogation Problem	15/05/2024	15/05/2024
									A072202	Low Impedance	15/05/2024	15/05/2024
									A090805	Incorrect Measurement	15/05/2024	15/05/2024
							F1901	Additional Surgery	A070909	Oversensing	15/05/2024	15/05/2024
									A0711	Interrogation Problem	15/05/2024	15/05/2024
									A072202	Low Impedance	15/05/2024	15/05/2024
									A090805	Incorrect Measurement	15/05/2024	15/05/2024
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A070909	Oversensing	15/05/2024	15/05/2024
									A0711	Interrogation Problem	15/05/2024	15/05/2024
									A072202	Low Impedance	15/05/2024	15/05/2024
									A090805	Incorrect Measurement	15/05/2024	15/05/2024
							F1901	Additional Surgery	A070909	Oversensing	15/05/2024	15/05/2024
									A0711	Interrogation Problem	15/05/2024	15/05/2024
									A072202	Low Impedance	15/05/2024	15/05/2024
									A090805	Incorrect Measurement	15/05/2024	15/05/2024
15459	2024	All other reportable incidents	47265	Dual-chamber implantable pacemaker, rate-responsive	CT1685	Cardiac pacemakers	E2103	Device Overstimulation of Tissue	A0908	Incorrect, Inadequate or Imprecise Result or Readings	15/01/2024	15/01/2024

							F08	Hospitalization or Prolonged Hospitalization	A0908	Incorrect, Inadequate or Imprecise Result or Readings	15/01/2024	15/01/2024
							F1905	Device Revision or Replacement	A0908	Incorrect, Inadequate or Imprecise Result or Readings	15/01/2024	15/01/2024
					CT257	Implantable cardiac pacemakers	E2103	Device Overstimulation of Tissue	A0908	Incorrect, Inadequate or Imprecise Result or Readings	15/01/2024	15/01/2024
							F08	Hospitalization or Prolonged Hospitalization	A0908	Incorrect, Inadequate or Imprecise Result or Readings	15/01/2024	15/01/2024
							F1905	Device Revision or Replacement	A0908	Incorrect, Inadequate or Imprecise Result or Readings	15/01/2024	15/01/2024
					CT2866	Implantable cardiac electrical stimulators	E2103	Device Overstimulation of Tissue	A0908	Incorrect, Inadequate or Imprecise Result or Readings	15/01/2024	15/01/2024
							F08	Hospitalization or Prolonged Hospitalization	A0908	Incorrect, Inadequate or Imprecise Result or Readings	15/01/2024	15/01/2024
							F1905	Device Revision or Replacement	A0908	Incorrect, Inadequate or Imprecise Result or Readings	15/01/2024	15/01/2024
15460	2024	Unanticipated serious deterioration in state of health	47265	Dual-chamber implantable pacemaker, rate-responsive	CT1685	Cardiac pacemakers	E0133	Stroke/CVA	A16	Protective Measures Problem		
							E060103	Atrial Flutter	A16	Protective Measures Problem		
							F08	Hospitalization or Prolonged Hospitalization	A16	Protective Measures Problem		
							F12	Serious Injury/ Illness/ Impairment	A16	Protective Measures Problem		
					CT257	Implantable cardiac pacemakers	E0133	Stroke/CVA	A16	Protective Measures Problem		
							E060103	Atrial Flutter	A16	Protective Measures Problem		
							F08	Hospitalization or Prolonged Hospitalization	A16	Protective Measures Problem		
							F12	Serious Injury/ Illness/ Impairment	A16	Protective Measures Problem		
					CT2866	Implantable cardiac electrical stimulators	E0133	Stroke/CVA	A16	Protective Measures Problem		
							E060103	Atrial Flutter	A16	Protective Measures Problem		
							F08	Hospitalization or Prolonged Hospitalization	A16	Protective Measures Problem		
							F12	Serious Injury/ Illness/ Impairment	A16	Protective Measures Problem		
15461	2024	All other reportable incidents	35852	Single-chamber implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A0709	Device Sensing Problem		
							F26	No Health Consequences or Impact	A0709	Device Sensing Problem		
15462	2024		47265	Dual-chamber implantable pacemaker, rate-responsive	CT1685	Cardiac pacemakers	F2301	Additional Device Required	A0705	Battery Problem		
					CT257	Implantable cardiac pacemakers	F2301	Additional Device Required	A0705	Battery Problem		
					CT2866	Implantable cardiac electrical stimulators	F2301	Additional Device Required	A0705	Battery Problem		
15463	2024	All other reportable incidents	17882	Physiologic-monitoring defibrillation system	CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A150105	Key or Button Unresponsive/not Working		
							F27	Problem Identified During Non-Clinical Procedure	A150105	Key or Button Unresponsive/not Working		
15464	2024	All other reportable incidents	36106	Implantable pacemaker analyser, intraoperative	CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A09	Output Problem		
									A13	Communication or Transmission Problem		
							F26	No Health Consequences or Impact	A09	Output Problem		
									A13	Communication or Transmission Problem		
15465	2024	All other reportable incidents	47265	Dual-chamber implantable pacemaker, rate-responsive	CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A23	Use of Device Problem		
							F26	No Health Consequences or Impact	A23	Use of Device Problem		
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A23	Use of Device Problem		
							F26	No Health Consequences or Impact	A23	Use of Device Problem		
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A23	Use of Device Problem		
							F26	No Health Consequences or Impact	A23	Use of Device Problem		
15466	2024	All other reportable incidents	47265	Dual-chamber implantable pacemaker, rate-responsive	CT1685	Cardiac pacemakers	E0601	Arrhythmia	A0712	Pacing Problem	11/04/2024	11/04/2024
							E060101	Asystole	A0712	Pacing Problem	11/04/2024	11/04/2024
							E0618	Non Specific EKG/ECG Changes	A0712	Pacing Problem	11/04/2024	11/04/2024
							F08	Hospitalization or Prolonged Hospitalization	A0712	Pacing Problem	11/04/2024	11/04/2024
							F12	Serious Injury/ Illness/ Impairment	A0712	Pacing Problem	11/04/2024	11/04/2024
							F1905	Device Revision or Replacement	A0712	Pacing Problem	11/04/2024	11/04/2024
							F23	Unexpected Medical Intervention	A0712	Pacing Problem	11/04/2024	11/04/2024
					CT257	Implantable cardiac pacemakers	E0601	Arrhythmia	A0712	Pacing Problem	11/04/2024	11/04/2024
							E060101	Asystole	A0712	Pacing Problem	11/04/2024	11/04/2024
							E0618	Non Specific EKG/ECG Changes	A0712	Pacing Problem	11/04/2024	11/04/2024
							F08	Hospitalization or Prolonged Hospitalization	A0712	Pacing Problem	11/04/2024	11/04/2024
							F12	Serious Injury/ Illness/ Impairment	A0712	Pacing Problem	11/04/2024	11/04/2024
							F1905	Device Revision or Replacement	A0712	Pacing Problem	11/04/2024	11/04/2024
							F23	Unexpected Medical Intervention	A0712	Pacing Problem	11/04/2024	11/04/2024
					CT2866	Implantable cardiac electrical stimulators	E0601	Arrhythmia	A0712	Pacing Problem	11/04/2024	11/04/2024
							E060101	Asystole	A0712	Pacing Problem	11/04/2024	11/04/2024
							E0618	Non Specific EKG/ECG Changes	A0712	Pacing Problem	11/04/2024	11/04/2024
							F08	Hospitalization or Prolonged Hospitalization	A0712	Pacing Problem	11/04/2024	11/04/2024
							F12	Serious Injury/ Illness/ Impairment	A0712	Pacing Problem	11/04/2024	11/04/2024
							F1905	Device Revision or Replacement	A0712	Pacing Problem	11/04/2024	11/04/2024
							F23	Unexpected Medical Intervention	A0712	Pacing Problem	11/04/2024	11/04/2024
15467	2024	Death	37265	Dual-chamber implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E0601	Arrhythmia	A071201	Failure to Convert Rhythm	26/10/2023	26/10/2023
							F02	Death	A071201	Failure to Convert Rhythm	26/10/2023	26/10/2023
15468	2024	All other reportable incidents	17882	Physiologic-monitoring defibrillation system	CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A0713	Defibrillation/Stimulation Problem		
							F26	No Health Consequences or Impact	A0713	Defibrillation/Stimulation Problem		
15469	2024	All other reportable incidents	47263	Cardiac resynchronization therapy implantable pacemaker	CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A0908	Incorrect, Inadequate or Imprecise Result or Readings	28/06/2024	28/06/2024
							F08	Hospitalization or Prolonged Hospitalization	A0908	Incorrect, Inadequate or Imprecise Result or Readings	28/06/2024	28/06/2024
							F1905	Device Revision or Replacement	A0908	Incorrect, Inadequate or Imprecise Result or Readings	28/06/2024	28/06/2024
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A0908	Incorrect, Inadequate or Imprecise Result or Readings	28/06/2024	28/06/2024
							F08	Hospitalization or Prolonged Hospitalization	A0908	Incorrect, Inadequate or Imprecise Result or Readings	28/06/2024	28/06/2024
							F1905	Device Revision or Replacement	A0908	Incorrect, Inadequate or Imprecise Result or Readings	28/06/2024	28/06/2024
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A0908	Incorrect, Inadequate or Imprecise Result or Readings	28/06/2024	28/06/2024
							F08	Hospitalization or Prolonged Hospitalization	A0908	Incorrect, Inadequate or Imprecise Result or Readings	28/06/2024	28/06/2024
							F1905	Device Revision or Replacement	A0908	Incorrect, Inadequate or Imprecise Result or Readings	28/06/2024	28/06/2024
15470	2024	All other reportable incidents	35852	Single-chamber implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E011903	Syncope/Fainting	A071201	Failure to Convert Rhythm		
							E0601	Arrhythmia	A071201	Failure to Convert Rhythm		
							E2402	Appropriate Clinical Signs, Symptoms and Conditions Term/Code not Available	A071201	Failure to Convert Rhythm		

							F12	Serious Injury/ Illness/ Impairment	A071201	Failure to Convert Rhythm		
15471	2024	All other reportable incidents	47263	Cardiac resynchronization therapy implantable pacemaker	CT1684	Cardiac pacemakers and associated devices	E2103	Device Overstimulation of Tissue	A071209	Extracardiac Stimulation	01/06/2024	01/06/2024
									A0908	Incorrect, Inadequate or Imprecise Result or Readings	01/06/2024	01/06/2024
							F08	Hospitalization or Prolonged Hospitalization	A071209	Extracardiac Stimulation	01/06/2024	01/06/2024
									A0908	Incorrect, Inadequate or Imprecise Result or Readings	01/06/2024	01/06/2024
							F1905	Device Revision or Replacement	A071209	Extracardiac Stimulation	01/06/2024	01/06/2024
									A0908	Incorrect, Inadequate or Imprecise Result or Readings	01/06/2024	01/06/2024
					CT1685	Cardiac pacemakers	E2103	Device Overstimulation of Tissue	A071209	Extracardiac Stimulation	01/06/2024	01/06/2024
									A0908	Incorrect, Inadequate or Imprecise Result or Readings	01/06/2024	01/06/2024
							F08	Hospitalization or Prolonged Hospitalization	A071209	Extracardiac Stimulation	01/06/2024	01/06/2024
									A0908	Incorrect, Inadequate or Imprecise Result or Readings	01/06/2024	01/06/2024
							F1905	Device Revision or Replacement	A071209	Extracardiac Stimulation	01/06/2024	01/06/2024
									A0908	Incorrect, Inadequate or Imprecise Result or Readings	01/06/2024	01/06/2024
					CT257	Implantable cardiac pacemakers	E2103	Device Overstimulation of Tissue	A071209	Extracardiac Stimulation	01/06/2024	01/06/2024
									A0908	Incorrect, Inadequate or Imprecise Result or Readings	01/06/2024	01/06/2024
							F08	Hospitalization or Prolonged Hospitalization	A071209	Extracardiac Stimulation	01/06/2024	01/06/2024
									A0908	Incorrect, Inadequate or Imprecise Result or Readings	01/06/2024	01/06/2024
							F1905	Device Revision or Replacement	A071209	Extracardiac Stimulation	01/06/2024	01/06/2024
									A0908	Incorrect, Inadequate or Imprecise Result or Readings	01/06/2024	01/06/2024
					CT2866	Implantable cardiac electrical stimulators	E2103	Device Overstimulation of Tissue	A071209	Extracardiac Stimulation	01/06/2024	01/06/2024
									A0908	Incorrect, Inadequate or Imprecise Result or Readings	01/06/2024	01/06/2024
							F08	Hospitalization or Prolonged Hospitalization	A071209	Extracardiac Stimulation	01/06/2024	01/06/2024
									A0908	Incorrect, Inadequate or Imprecise Result or Readings	01/06/2024	01/06/2024
							F1905	Device Revision or Replacement	A071209	Extracardiac Stimulation	01/06/2024	01/06/2024
									A0908	Incorrect, Inadequate or Imprecise Result or Readings	01/06/2024	01/06/2024
15472	2024	All other reportable incidents	47205	Cardiac pulse generator programmer	CT1684	Cardiac pacemakers and associated devices	E060109	Tachycardia	A0902	Display or Visual Feedback Problem		
									A11	Computer Software Problem		
							F1203	Life Threatening Illness or Injury	A0902	Display or Visual Feedback Problem		
									A11	Computer Software Problem		
							F23	Unexpected Medical Intervention	A0902	Display or Visual Feedback Problem		
									A11	Computer Software Problem		
15473	2024	All other reportable incidents	47265	Dual-chamber implantable pacemaker, rate-responsive	CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem	17/06/2024	17/06/2024
									A070504	Premature Discharge of Battery	17/06/2024	17/06/2024
							F08	Hospitalization or Prolonged Hospitalization	A0705	Battery Problem	17/06/2024	17/06/2024
									A070504	Premature Discharge of Battery	17/06/2024	17/06/2024
							F1905	Device Revision or Replacement	A0705	Battery Problem	17/06/2024	17/06/2024
									A070504	Premature Discharge of Battery	17/06/2024	17/06/2024
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem	17/06/2024	17/06/2024
									A070504	Premature Discharge of Battery	17/06/2024	17/06/2024
							F08	Hospitalization or Prolonged Hospitalization	A0705	Battery Problem	17/06/2024	17/06/2024
									A070504	Premature Discharge of Battery	17/06/2024	17/06/2024
							F1905	Device Revision or Replacement	A0705	Battery Problem	17/06/2024	17/06/2024
									A070504	Premature Discharge of Battery	17/06/2024	17/06/2024
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem	17/06/2024	17/06/2024
									A070504	Premature Discharge of Battery	17/06/2024	17/06/2024
							F08	Hospitalization or Prolonged Hospitalization	A0705	Battery Problem	17/06/2024	17/06/2024
									A070504	Premature Discharge of Battery	17/06/2024	17/06/2024
							F1905	Device Revision or Replacement	A0705	Battery Problem	17/06/2024	17/06/2024
									A070504	Premature Discharge of Battery	17/06/2024	17/06/2024
15474	2024	All other reportable incidents	47265	Dual-chamber implantable pacemaker, rate-responsive	CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
15475	2024	All other reportable incidents	47265	Dual-chamber implantable pacemaker, rate-responsive	CT1685	Cardiac pacemakers	E2312	Fatigue	A0908	Incorrect, Inadequate or Imprecise Result or Readings	28/06/2024	28/06/2024
							F08	Hospitalization or Prolonged Hospitalization	A0908	Incorrect, Inadequate or Imprecise Result or Readings	28/06/2024	28/06/2024
							F1905	Device Revision or Replacement	A0908	Incorrect, Inadequate or Imprecise Result or Readings	28/06/2024	28/06/2024
					CT257	Implantable cardiac pacemakers	E2312	Fatigue	A0908	Incorrect, Inadequate or Imprecise Result or Readings	28/06/2024	28/06/2024
							F08	Hospitalization or Prolonged Hospitalization	A0908	Incorrect, Inadequate or Imprecise Result or Readings	28/06/2024	28/06/2024
							F1905	Device Revision or Replacement	A0908	Incorrect, Inadequate or Imprecise Result or Readings	28/06/2024	28/06/2024
					CT2866	Implantable cardiac electrical stimulators	E2312	Fatigue	A0908	Incorrect, Inadequate or Imprecise Result or Readings	28/06/2024	28/06/2024
							F08	Hospitalization or Prolonged Hospitalization	A0908	Incorrect, Inadequate or Imprecise Result or Readings	28/06/2024	28/06/2024
							F1905	Device Revision or Replacement	A0908	Incorrect, Inadequate or Imprecise Result or Readings	28/06/2024	28/06/2024
15476	2024	All other reportable incidents	47265	Dual-chamber implantable pacemaker, rate-responsive	CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	23/08/2023	23/08/2023
									A160602	Premature End-of-Life Indicator	23/08/2023	23/08/2023
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	23/08/2023	23/08/2023
									A160602	Premature End-of-Life Indicator	23/08/2023	23/08/2023
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	23/08/2023	23/08/2023
									A160602	Premature End-of-Life Indicator	23/08/2023	23/08/2023
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	23/08/2023	23/08/2023
									A160602	Premature End-of-Life Indicator	23/08/2023	23/08/2023

							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	23/08/2023	23/08/2023
									A160602	Premature End-of-Life Indicator	23/08/2023	23/08/2023
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	23/08/2023	23/08/2023
									A160602	Premature End-of-Life Indicator	23/08/2023	23/08/2023
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	23/08/2023	23/08/2023
									A160602	Premature End-of-Life Indicator	23/08/2023	23/08/2023
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	23/08/2023	23/08/2023
									A160602	Premature End-of-Life Indicator	23/08/2023	23/08/2023
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	23/08/2023	23/08/2023
									A160602	Premature End-of-Life Indicator	23/08/2023	23/08/2023
15477	2024		47265	Dual-chamber implantable pacemaker, rate-responsive	CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A0201	Product Quality Problem		
									A23	Use of Device Problem		
							F26	No Health Consequences or Impact	A0201	Product Quality Problem		
									A23	Use of Device Problem		
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A0201	Product Quality Problem		
									A23	Use of Device Problem		
							F26	No Health Consequences or Impact	A0201	Product Quality Problem		
									A23	Use of Device Problem		
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A0201	Product Quality Problem		
									A23	Use of Device Problem		
							F26	No Health Consequences or Impact	A0201	Product Quality Problem		
									A23	Use of Device Problem		
15478	2024	All other reportable incidents	47265	Dual-chamber implantable pacemaker, rate-responsive	CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	14/04/2024	14/04/2024
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery	14/04/2024	14/04/2024
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	14/04/2024	14/04/2024
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery	14/04/2024	14/04/2024
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	14/04/2024	14/04/2024
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery	14/04/2024	14/04/2024
15479	2024	All other reportable incidents	35852	Single-chamber implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A070909	Oversensing		
									A072001	Electromagnetic Interference		
									A090801	Signal Artifact/Noise		
							F12	Serious Injury/ Illness/ Impairment	A070909	Oversensing		
									A072001	Electromagnetic Interference		
									A090801	Signal Artifact/Noise		
							F2307	Reprogramming of Device	A070909	Oversensing		
									A072001	Electromagnetic Interference		
									A090801	Signal Artifact/Noise		
15480	2024	All other reportable incidents	35852	Single-chamber implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A0508	Noise, Audible		
									A070102	High Capture Threshold		
									A070504	Premature Discharge of Battery		
							F24	Insufficient Health Impact Information	A0508	Noise, Audible		
									A070102	High Capture Threshold		
									A070504	Premature Discharge of Battery		
15481	2024	All other reportable incidents	17882	Physiologic-monitoring defibrillation system	CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A090805	Incorrect Measurement		
							F27	Problem Identified During Non-Clinical Procedure	A090805	Incorrect Measurement		
15482	2024	All other reportable incidents	47265	Dual-chamber implantable pacemaker, rate-responsive	CT1685	Cardiac pacemakers	E0119	Loss of Consciousness	A0712	Pacing Problem	29/06/2024	29/06/2024
							E0618	Non Specific EKG/ECG Changes	A0712	Pacing Problem	29/06/2024	29/06/2024
							F1203	Life Threatening Illness or Injury	A0712	Pacing Problem	29/06/2024	29/06/2024
					CT257	Implantable cardiac pacemakers	E0119	Loss of Consciousness	A0712	Pacing Problem	29/06/2024	29/06/2024
							E0618	Non Specific EKG/ECG Changes	A0712	Pacing Problem	29/06/2024	29/06/2024
							F1203	Life Threatening Illness or Injury	A0712	Pacing Problem	29/06/2024	29/06/2024
					CT2866	Implantable cardiac electrical stimulators	E0119	Loss of Consciousness	A0712	Pacing Problem	29/06/2024	29/06/2024
							E0618	Non Specific EKG/ECG Changes	A0712	Pacing Problem	29/06/2024	29/06/2024
							F1203	Life Threatening Illness or Injury	A0712	Pacing Problem	29/06/2024	29/06/2024
15483	2024	All other reportable incidents	60789	Intracardiac pacemaker	CT1685	Cardiac pacemakers	E0137	Transient Ischemic Attack	A27	Appropriate Device Problem Term/Code not Available		
							E0724	Hyperventilation	A27	Appropriate Device Problem Term/Code not Available		
							E1621	Muscle Weakness/Atrophy	A27	Appropriate Device Problem Term/Code not Available		
							E233001	Chest Pain	A27	Appropriate Device Problem Term/Code not Available		
							F08	Hospitalization or Prolonged Hospitalization	A27	Appropriate Device Problem Term/Code not Available		
							F12	Serious Injury/ Illness/ Impairment	A27	Appropriate Device Problem Term/Code not Available		
							F2303	Medication Required	A27	Appropriate Device Problem Term/Code not Available		
					CT257	Implantable cardiac pacemakers	E0137	Transient Ischemic Attack	A27	Appropriate Device Problem Term/Code not Available		
							E0724	Hyperventilation	A27	Appropriate Device Problem Term/Code not Available		
							E1621	Muscle Weakness/Atrophy	A27	Appropriate Device Problem Term/Code not Available		
							E233001	Chest Pain	A27	Appropriate Device Problem Term/Code not Available		
							F08	Hospitalization or Prolonged Hospitalization	A27	Appropriate Device Problem Term/Code not Available		
							F12	Serious Injury/ Illness/ Impairment	A27	Appropriate Device Problem Term/Code not Available		
							F2303	Medication Required	A27	Appropriate Device Problem Term/Code not Available		
					CT2818	Catheterization/Cannulation/Fluid path devices	E0137	Transient Ischemic Attack	A27	Appropriate Device Problem Term/Code not Available		
							E0724	Hyperventilation	A27	Appropriate Device Problem Term/Code not Available		
							E1621	Muscle Weakness/Atrophy	A27	Appropriate Device Problem Term/Code not Available		
							E233001	Chest Pain	A27	Appropriate Device Problem Term/Code not Available		
							F08	Hospitalization or Prolonged Hospitalization	A27	Appropriate Device Problem Term/Code not Available		

							F12	Serious Injury/ Illness/ Impairment	A27	Appropriate Device Problem Term/Code not Available		
							F2303	Medication Required	A27	Appropriate Device Problem Term/Code not Available		
					CT2866	Implantable cardiac electrical stimulators	E0137	Transient Ischemic Attack	A27	Appropriate Device Problem Term/Code not Available		
							E0724	Hyperventilation	A27	Appropriate Device Problem Term/Code not Available		
							E1621	Muscle Weakness/Atrophy	A27	Appropriate Device Problem Term/Code not Available		
							E233001	Chest Pain	A27	Appropriate Device Problem Term/Code not Available		
							F08	Hospitalization or Prolonged Hospitalization	A27	Appropriate Device Problem Term/Code not Available		
							F12	Serious Injury/ Illness/ Impairment	A27	Appropriate Device Problem Term/Code not Available		
							F2303	Medication Required	A27	Appropriate Device Problem Term/Code not Available		
15484	2024	All other reportable incidents	17882	Physiologic-monitoring defibrillation system	CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070603	Failure to Charge		
							F27	Problem Identified During Non-Clinical Procedure	A070603	Failure to Charge		
15485	2024	All other reportable incidents	35852	Single-chamber implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
15486	2024	All other reportable incidents	37265	Dual-chamber implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E210401	Shock from Patient Lead(s)	A0709	Device Sensing Problem		
									A071302	Inappropriate/Inadequate Shock/Stimulation		
							F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A0709	Device Sensing Problem		
									A071302	Inappropriate/Inadequate Shock/Stimulation		
							F1203	Life Threatening Illness or Injury	A0709	Device Sensing Problem		
									A071302	Inappropriate/Inadequate Shock/Stimulation		
15487	2024		47265	Dual-chamber implantable pacemaker, rate-responsive	CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem		
							F26	No Health Consequences or Impact	A0705	Battery Problem		
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem		
							F26	No Health Consequences or Impact	A0705	Battery Problem		
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem		
							F26	No Health Consequences or Impact	A0705	Battery Problem		
15488	2024	All other reportable incidents	17882	Physiologic-monitoring defibrillation system	CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A0719	Unexpected Shutdown		
							F26	No Health Consequences or Impact	A0719	Unexpected Shutdown		
15489	2024	All other reportable incidents	17882	Physiologic-monitoring defibrillation system	CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A0709	Device Sensing Problem		
							F26	No Health Consequences or Impact	A0709	Device Sensing Problem		
15490	2024	All other reportable incidents	47270	Cardiac resynchronization therapy implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	04/02/2025	04/02/2025
									A0908	Incorrect, Inadequate or Imprecise Result or Readings	04/02/2025	04/02/2025
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	04/02/2025	04/02/2025
									A0908	Incorrect, Inadequate or Imprecise Result or Readings	04/02/2025	04/02/2025
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	04/02/2025	04/02/2025
									A0908	Incorrect, Inadequate or Imprecise Result or Readings	04/02/2025	04/02/2025
15491	2024	All other reportable incidents	17882	Physiologic-monitoring defibrillation system	CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A0719	Unexpected Shutdown		
							F26	No Health Consequences or Impact	A0719	Unexpected Shutdown		
15492	2024	All other reportable incidents	47263	Cardiac resynchronization therapy implantable pacemaker	CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A160105	False Alarm		
							F26	No Health Consequences or Impact	A160105	False Alarm		
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A160105	False Alarm		
							F26	No Health Consequences or Impact	A160105	False Alarm		
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A160105	False Alarm		
							F26	No Health Consequences or Impact	A160105	False Alarm		
15493	2024	All other reportable incidents	47265	Dual-chamber implantable pacemaker, rate-responsive	CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	17/06/2024	17/06/2024
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	17/06/2024	17/06/2024
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	17/06/2024	17/06/2024
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	17/06/2024	17/06/2024
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	17/06/2024	17/06/2024
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	17/06/2024	17/06/2024
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	17/06/2024	17/06/2024
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	17/06/2024	17/06/2024
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	17/06/2024	17/06/2024
15494	2024	All other reportable incidents	47265	Dual-chamber implantable pacemaker, rate-responsive	CT1685	Cardiac pacemakers	E0619	Pericardial Effusion	A0712	Pacing Problem		
							E2403	No Clinical Signs, Symptoms or Conditions	A0712	Pacing Problem		
							F08	Hospitalization or Prolonged Hospitalization	A0712	Pacing Problem		
							F12	Serious Injury/ Illness/ Impairment	A0712	Pacing Problem		
							F1901	Additional Surgery	A0712	Pacing Problem		
							F2307	Reprogramming of Device	A0712	Pacing Problem		
					CT257	Implantable cardiac pacemakers	E0619	Pericardial Effusion	A0712	Pacing Problem		
							E2403	No Clinical Signs, Symptoms or Conditions	A0712	Pacing Problem		
							F08	Hospitalization or Prolonged Hospitalization	A0712	Pacing Problem		
							F12	Serious Injury/ Illness/ Impairment	A0712	Pacing Problem		
							F1901	Additional Surgery	A0712	Pacing Problem		
							F2307	Reprogramming of Device	A0712	Pacing Problem		
					CT2866	Implantable cardiac electrical stimulators	E0619	Pericardial Effusion	A0712	Pacing Problem		
							E2403	No Clinical Signs, Symptoms or Conditions	A0712	Pacing Problem		
							F08	Hospitalization or Prolonged Hospitalization	A0712	Pacing Problem		
							F12	Serious Injury/ Illness/ Impairment	A0712	Pacing Problem		
							F1901	Additional Surgery	A0712	Pacing Problem		
							F2307	Reprogramming of Device	A0712	Pacing Problem		
15495	2024	All other reportable incidents	47265	Dual-chamber implantable pacemaker, rate-responsive	CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	20/02/2024	20/02/2024
							F08	Hospitalization or Prolonged Hospitalization	A24	No Apparent Device Problem	20/02/2024	20/02/2024
							F1905	Device Revision or Replacement	A24	No Apparent Device Problem	20/02/2024	20/02/2024
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	20/02/2024	20/02/2024
							F08	Hospitalization or Prolonged Hospitalization	A24	No Apparent Device Problem	20/02/2024	20/02/2024
							F1905	Device Revision or Replacement	A24	No Apparent Device Problem	20/02/2024	20/02/2024
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	20/02/2024	20/02/2024
							F08	Hospitalization or Prolonged Hospitalization	A24	No Apparent Device Problem	20/02/2024	20/02/2024
							F1905	Device Revision or Replacement	A24	No Apparent Device Problem	20/02/2024	20/02/2024

15496	2024	All other reportable incidents	37265	Dual-chamber implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E011902	Presyncope	A0709	Device Sensing Problem		
									A071301	Failure to Deliver Shock/Stimulation		
							E060109	Tachycardia	A0709	Device Sensing Problem		
									A071301	Failure to Deliver Shock/Stimulation		
							F08	Hospitalization or Prolonged Hospitalization	A0709	Device Sensing Problem		
									A071301	Failure to Deliver Shock/Stimulation		
							F1001	Absence of Treatment	A0709	Device Sensing Problem		
									A071301	Failure to Deliver Shock/Stimulation		
							F1203	Life Threatening Illness or Injury	A0709	Device Sensing Problem		
									A071301	Failure to Deliver Shock/Stimulation		
15497	2024	All other reportable incidents	17882	Physiologic-monitoring defibrillation system	CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A07	Electrical /Electronic Property Problem		
									A0719	Unexpected Shutdown		
							F26	No Health Consequences or Impact	A07	Electrical /Electronic Property Problem		
									A0719	Unexpected Shutdown		
15498	2024	All other reportable incidents	35852	Single-chamber implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A070101	Failure to Capture		
									A070102	High Capture Threshold		
							F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A070101	Failure to Capture		
									A070102	High Capture Threshold		
15499	2024	All other reportable incidents	47263	Cardiac resynchronization therapy implantable pacemaker	CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	04/07/2024	04/07/2024
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	04/07/2024	04/07/2024
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	04/07/2024	04/07/2024
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	04/07/2024	04/07/2024
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	04/07/2024	04/07/2024
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	04/07/2024	04/07/2024
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	04/07/2024	04/07/2024
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	04/07/2024	04/07/2024
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	04/07/2024	04/07/2024
15500	2024	All other reportable incidents	47265	Dual-chamber implantable pacemaker, rate-responsive	CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A05	Mechanical Problem	28/04/2024	28/04/2024
							F26	No Health Consequences or Impact	A05	Mechanical Problem	28/04/2024	28/04/2024
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A05	Mechanical Problem	28/04/2024	28/04/2024
							F26	No Health Consequences or Impact	A05	Mechanical Problem	28/04/2024	28/04/2024
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A05	Mechanical Problem	28/04/2024	28/04/2024
							F26	No Health Consequences or Impact	A05	Mechanical Problem	28/04/2024	28/04/2024
15501	2024		60789	Intracardiac pacemaker	CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A0508	Noise, Audible		
									A1605	Reset Problem		
							F26	No Health Consequences or Impact	A0508	Noise, Audible		
									A1605	Reset Problem		
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A0508	Noise, Audible		
									A1605	Reset Problem		
							F26	No Health Consequences or Impact	A0508	Noise, Audible		
									A1605	Reset Problem		
					CT2818	Catheterization/Cannulation/Fluid path devices	E2403	No Clinical Signs, Symptoms or Conditions	A0508	Noise, Audible		
									A1605	Reset Problem		
							F26	No Health Consequences or Impact	A0508	Noise, Audible		
									A1605	Reset Problem		
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A0508	Noise, Audible		
									A1605	Reset Problem		
							F26	No Health Consequences or Impact	A0508	Noise, Audible		
									A1605	Reset Problem		
15502	2024		35852	Single-chamber implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A25	No Apparent Adverse Event		
							F26	No Health Consequences or Impact	A25	No Apparent Adverse Event		
15503	2024	All other reportable incidents	37265	Dual-chamber implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem	02/10/2024	02/10/2024
									A070504	Premature Discharge of Battery	02/10/2024	02/10/2024
							F08	Hospitalization or Prolonged Hospitalization	A0705	Battery Problem	02/10/2024	02/10/2024
									A070504	Premature Discharge of Battery	02/10/2024	02/10/2024
							F1903	Device Explantation	A0705	Battery Problem	02/10/2024	02/10/2024
									A070504	Premature Discharge of Battery	02/10/2024	02/10/2024
15504	2024	All other reportable incidents	47265	Dual-chamber implantable pacemaker, rate-responsive	CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
							F24	Insufficient Health Impact Information	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
							F24	Insufficient Health Impact Information	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
							F24	Insufficient Health Impact Information	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
15505	2024	All other reportable incidents	17882	Physiologic-monitoring defibrillation system	CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A0801	Failure to Calibrate		
							F27	Problem Identified During Non-Clinical Procedure	A0801	Failure to Calibrate		
15506	2024	All other reportable incidents	35852	Single-chamber implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	09/12/2024	09/12/2024
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	09/12/2024	09/12/2024
							F1903	Device Explantation	A070504	Premature Discharge of Battery	09/12/2024	09/12/2024
15507	2024	All other reportable incidents	47265	Dual-chamber implantable pacemaker, rate-responsive	CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem	15/02/2024	15/02/2024
									A070504	Premature Discharge of Battery	15/02/2024	15/02/2024
							F08	Hospitalization or Prolonged Hospitalization	A0705	Battery Problem	15/02/2024	15/02/2024
									A070504	Premature Discharge of Battery	15/02/2024	15/02/2024
							F1905	Device Revision or Replacement	A0705	Battery Problem	15/02/2024	15/02/2024
									A070504	Premature Discharge of Battery	15/02/2024	15/02/2024
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem	15/02/2024	15/02/2024
									A070504	Premature Discharge of Battery	15/02/2024	15/02/2024
							F08	Hospitalization or Prolonged Hospitalization	A0705	Battery Problem	15/02/2024	15/02/2024
									A070504	Premature Discharge of Battery	15/02/2024	15/02/2024
							F1905	Device Revision or Replacement	A0705	Battery Problem	15/02/2024	15/02/2024
									A070504	Premature Discharge of Battery	15/02/2024	15/02/2024
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem	15/02/2024	15/02/2024
									A070504	Premature Discharge of Battery	15/02/2024	15/02/2024

							F08	Hospitalization or Prolonged Hospitalization	A0705	Battery Problem	15/02/2024	15/02/2024
									A070504	Premature Discharge of Battery	15/02/2024	15/02/2024
							F1905	Device Revision or Replacement	A0705	Battery Problem	15/02/2024	15/02/2024
									A070504	Premature Discharge of Battery	15/02/2024	15/02/2024
15508	2024	All other reportable incidents	35852	Single-chamber implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A0709	Device Sensing Problem		
									A070909	Oversensing		
									A1301	Failure to Read Input Signal		
							F13	Misdiagnosis/ Misclassification	A0709	Device Sensing Problem		
									A070909	Oversensing		
									A1301	Failure to Read Input Signal		
							F2203	Imaging Required	A0709	Device Sensing Problem		
									A070909	Oversensing		
									A1301	Failure to Read Input Signal		
15509	2024	All other reportable incidents	47265	Dual-chamber implantable pacemaker, rate-responsive	CT1685	Cardiac pacemakers	E0601	Arrhythmia	A070504	Premature Discharge of Battery	02/07/2024	02/07/2024
									A071102	Failure to Interrogate	02/07/2024	02/07/2024
									A071204	No Pacing	02/07/2024	02/07/2024
									A0723	Unexpected Mode Switch	02/07/2024	02/07/2024
							E0717	Dyspnea	A070504	Premature Discharge of Battery	02/07/2024	02/07/2024
									A071102	Failure to Interrogate	02/07/2024	02/07/2024
									A071204	No Pacing	02/07/2024	02/07/2024
									A0723	Unexpected Mode Switch	02/07/2024	02/07/2024
							E2312	Fatigue	A070504	Premature Discharge of Battery	02/07/2024	02/07/2024
									A071102	Failure to Interrogate	02/07/2024	02/07/2024
									A071204	No Pacing	02/07/2024	02/07/2024
									A0723	Unexpected Mode Switch	02/07/2024	02/07/2024
							F1901	Additional Surgery	A070504	Premature Discharge of Battery	02/07/2024	02/07/2024
									A071102	Failure to Interrogate	02/07/2024	02/07/2024
									A071204	No Pacing	02/07/2024	02/07/2024
									A0723	Unexpected Mode Switch	02/07/2024	02/07/2024
					CT257	Implantable cardiac pacemakers	E0601	Arrhythmia	A070504	Premature Discharge of Battery	02/07/2024	02/07/2024
									A071102	Failure to Interrogate	02/07/2024	02/07/2024
									A071204	No Pacing	02/07/2024	02/07/2024
									A0723	Unexpected Mode Switch	02/07/2024	02/07/2024
							E0717	Dyspnea	A070504	Premature Discharge of Battery	02/07/2024	02/07/2024
									A071102	Failure to Interrogate	02/07/2024	02/07/2024
									A071204	No Pacing	02/07/2024	02/07/2024
									A0723	Unexpected Mode Switch	02/07/2024	02/07/2024
							E2312	Fatigue	A070504	Premature Discharge of Battery	02/07/2024	02/07/2024
									A071102	Failure to Interrogate	02/07/2024	02/07/2024
									A071204	No Pacing	02/07/2024	02/07/2024
									A0723	Unexpected Mode Switch	02/07/2024	02/07/2024
							F1901	Additional Surgery	A070504	Premature Discharge of Battery	02/07/2024	02/07/2024
									A071102	Failure to Interrogate	02/07/2024	02/07/2024
									A071204	No Pacing	02/07/2024	02/07/2024
									A0723	Unexpected Mode Switch	02/07/2024	02/07/2024
					CT2866	Implantable cardiac electrical stimulators	E0601	Arrhythmia	A070504	Premature Discharge of Battery	02/07/2024	02/07/2024
									A071102	Failure to Interrogate	02/07/2024	02/07/2024
									A071204	No Pacing	02/07/2024	02/07/2024
									A0723	Unexpected Mode Switch	02/07/2024	02/07/2024
							E0717	Dyspnea	A070504	Premature Discharge of Battery	02/07/2024	02/07/2024
									A071102	Failure to Interrogate	02/07/2024	02/07/2024
									A071204	No Pacing	02/07/2024	02/07/2024
									A0723	Unexpected Mode Switch	02/07/2024	02/07/2024
							E2312	Fatigue	A070504	Premature Discharge of Battery	02/07/2024	02/07/2024
									A071102	Failure to Interrogate	02/07/2024	02/07/2024
									A071204	No Pacing	02/07/2024	02/07/2024
									A0723	Unexpected Mode Switch	02/07/2024	02/07/2024
							F1901	Additional Surgery	A070504	Premature Discharge of Battery	02/07/2024	02/07/2024
									A071102	Failure to Interrogate	02/07/2024	02/07/2024
									A071204	No Pacing	02/07/2024	02/07/2024
									A0723	Unexpected Mode Switch	02/07/2024	02/07/2024
15510	2024	All other reportable incidents			CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070910	Undersensing		
							F26	No Health Consequences or Impact	A070910	Undersensing		
15511	2024	All other reportable incidents	47265	Dual-chamber implantable pacemaker, rate-responsive	CT1685	Cardiac pacemakers	E0601	Arrhythmia	A070101	Failure to Capture		
									A071206	Pacing Asynchronously		
							F23	Unexpected Medical Intervention	A070101	Failure to Capture		
									A071206	Pacing Asynchronously		
					CT257	Implantable cardiac pacemakers	E0601	Arrhythmia	A070101	Failure to Capture		
									A071206	Pacing Asynchronously		
							F23	Unexpected Medical Intervention	A070101	Failure to Capture		
									A071206	Pacing Asynchronously		
					CT2866	Implantable cardiac electrical stimulators	E0601	Arrhythmia	A070101	Failure to Capture		
									A071206	Pacing Asynchronously		
							F23	Unexpected Medical Intervention	A070101	Failure to Capture		
									A071206	Pacing Asynchronously		
15512	2024	All other reportable incidents	17882	Physiologic-monitoring defibrillation system	CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A0514	Damaged Thread		
							F27	Problem Identified During Non-Clinical Procedure	A0514	Damaged Thread		
15513	2024	All other reportable incidents	17882	Physiologic-monitoring defibrillation system	CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A0514	Damaged Thread		
							F27	Problem Identified During Non-Clinical Procedure	A0514	Damaged Thread		
15514	2024	All other reportable incidents	17882	Physiologic-monitoring defibrillation system	CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A0514	Damaged Thread		
							F27	Problem Identified During Non-Clinical Procedure	A0514	Damaged Thread		
15515	2024	Unanticipated serious deterioration in state of health	47265	Dual-chamber implantable pacemaker, rate-responsive	CT1685	Cardiac pacemakers	E060101	Asystole	A0712	Pacing Problem	09/07/2024	09/07/2024
							F1203	Life Threatening Illness or Injury	A0712	Pacing Problem	09/07/2024	09/07/2024
					CT257	Implantable cardiac pacemakers	E060101	Asystole	A0712	Pacing Problem	09/07/2024	09/07/2024
							F1203	Life Threatening Illness or Injury	A0712	Pacing Problem	09/07/2024	09/07/2024
					CT2866	Implantable cardiac electrical stimulators	E060101	Asystole	A0712	Pacing Problem	09/07/2024	09/07/2024
							F1203	Life Threatening Illness or Injury	A0712	Pacing Problem	09/07/2024	09/07/2024
15516	2024	All other reportable incidents	17882	Physiologic-monitoring defibrillation system	CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A07	Electrical /Electronic Property Problem		

							F27	Problem Identified During Non-Clinical Procedure	A07	Electrical /Electronic Property Problem		
15517	2024	All other reportable incidents	35852	Single-chamber implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A0711	Interrogation Problem	08/07/2024	08/07/2024
									A1304	Telemetry Discrepancy	08/07/2024	08/07/2024
							F08	Hospitalization or Prolonged Hospitalization	A0711	Interrogation Problem	08/07/2024	08/07/2024
									A1304	Telemetry Discrepancy	08/07/2024	08/07/2024
							F1905	Device Revision or Replacement	A0711	Interrogation Problem	08/07/2024	08/07/2024
									A1304	Telemetry Discrepancy	08/07/2024	08/07/2024
15518	2024	All other reportable incidents	17882	Physiologic-monitoring defibrillation system	CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070803	Failure to Power Up		
							F27	Problem Identified During Non-Clinical Procedure	A070803	Failure to Power Up		
15519	2024		47265	Dual-chamber implantable pacemaker, rate-responsive	CT141	Electrodes/Leads and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem		
							F1901	Additional Surgery	A24	No Apparent Device Problem		
					CT145	Cardiovascular devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem		
							F1901	Additional Surgery	A24	No Apparent Device Problem		
					CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem		
							F1901	Additional Surgery	A24	No Apparent Device Problem		
					CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem		
							F1901	Additional Surgery	A24	No Apparent Device Problem		
					CT233	Surgical	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem		
							F1901	Additional Surgery	A24	No Apparent Device Problem		
					CT2406	Implantable	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem		
							F1901	Additional Surgery	A24	No Apparent Device Problem		
					CT2407	Active implantable	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem		
							F1901	Additional Surgery	A24	No Apparent Device Problem		
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem		
							F1901	Additional Surgery	A24	No Apparent Device Problem		
					CT2727	Active implant devices/ancillaries	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem		
							F1901	Additional Surgery	A24	No Apparent Device Problem		
					CT2750	Implantable cardiac electrical stimulators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem		
							F1901	Additional Surgery	A24	No Apparent Device Problem		
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem		
							F1901	Additional Surgery	A24	No Apparent Device Problem		
					CT2985	Active	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem		
							F1901	Additional Surgery	A24	No Apparent Device Problem		
					CT3001	Implantable metal	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem		
							F1901	Additional Surgery	A24	No Apparent Device Problem		
					CT3002	Implantable synthetic polymer	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem		
							F1901	Additional Surgery	A24	No Apparent Device Problem		
					CT3006	Implantable inorganic material	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem		
							F1901	Additional Surgery	A24	No Apparent Device Problem		
					CT314	Electrically powered	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem		
							F1901	Additional Surgery	A24	No Apparent Device Problem		
					CT321	Long-term surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem		
							F1901	Additional Surgery	A24	No Apparent Device Problem		
					CT449	Generators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem		
							F1901	Additional Surgery	A24	No Apparent Device Problem		
					CT981	Single use	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem		
							F1901	Additional Surgery	A24	No Apparent Device Problem		
					CT983	Surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem		
							F1901	Additional Surgery	A24	No Apparent Device Problem		
15520	2024	All other reportable incidents	17882	Physiologic-monitoring defibrillation system	CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070803	Failure to Power Up		
							F27	Problem Identified During Non-Clinical Procedure	A070803	Failure to Power Up		
15521	2024		47265	Dual-chamber implantable pacemaker, rate-responsive	CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem		
							F1901	Additional Surgery	A24	No Apparent Device Problem		
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem		
							F1901	Additional Surgery	A24	No Apparent Device Problem		
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem		
							F1901	Additional Surgery	A24	No Apparent Device Problem		
15522	2024	All other reportable incidents			CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A04	Material Integrity Problem	24/06/2024	24/06/2024
									A072201	High Impedance	24/06/2024	24/06/2024
							F08	Hospitalization or Prolonged Hospitalization	A04	Material Integrity Problem	24/06/2024	24/06/2024
									A072201	High Impedance	24/06/2024	24/06/2024
							F1905	Device Revision or Replacement	A04	Material Integrity Problem	24/06/2024	24/06/2024
									A072201	High Impedance	24/06/2024	24/06/2024
15523	2024	All other reportable incidents	47263	Cardiac resynchronization therapy implantable pacemaker	CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A0908	Incorrect, Inadequate or Imprecise Result or Readings	04/07/2024	04/07/2024
							F08	Hospitalization or Prolonged Hospitalization	A0908	Incorrect, Inadequate or Imprecise Result or Readings	04/07/2024	04/07/2024
							F1905	Device Revision or Replacement	A0908	Incorrect, Inadequate or Imprecise Result or Readings	04/07/2024	04/07/2024
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A0908	Incorrect, Inadequate or Imprecise Result or Readings	04/07/2024	04/07/2024
							F08	Hospitalization or Prolonged Hospitalization	A0908	Incorrect, Inadequate or Imprecise Result or Readings	04/07/2024	04/07/2024
							F1905	Device Revision or Replacement	A0908	Incorrect, Inadequate or Imprecise Result or Readings	04/07/2024	04/07/2024
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A0908	Incorrect, Inadequate or Imprecise Result or Readings	04/07/2024	04/07/2024
							F08	Hospitalization or Prolonged Hospitalization	A0908	Incorrect, Inadequate or Imprecise Result or Readings	04/07/2024	04/07/2024
							F1905	Device Revision or Replacement	A0908	Incorrect, Inadequate or Imprecise Result or Readings	04/07/2024	04/07/2024
15524	2024	All other reportable incidents	47270	Cardiac resynchronization therapy implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A0508	Noise, Audible	28/11/2024	28/11/2024
									A0705	Battery Problem	28/11/2024	28/11/2024
									A070504	Premature Discharge of Battery	28/11/2024	28/11/2024
									A0908	Incorrect, Inadequate or Imprecise Result or Readings	28/11/2024	28/11/2024
							F08	Hospitalization or Prolonged Hospitalization	A0508	Noise, Audible	28/11/2024	28/11/2024
									A0705	Battery Problem	28/11/2024	28/11/2024
									A070504	Premature Discharge of Battery	28/11/2024	28/11/2024

								A0908	Incorrect, Inadequate or Imprecise Result or Readings	28/11/2024	28/11/2024
						F1905	Device Revision or Replacement	A0508	Noise, Audible	28/11/2024	28/11/2024
								A0705	Battery Problem	28/11/2024	28/11/2024
								A070504	Premature Discharge of Battery	28/11/2024	28/11/2024
								A0908	Incorrect, Inadequate or Imprecise Result or Readings	28/11/2024	28/11/2024
15525	2024	All other reportable incidents	47265	Dual-chamber implantable pacemaker, rate-responsive	CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem	10/07/2024	10/07/2024
									A070504	Premature Discharge of Battery	10/07/2024	10/07/2024
									A0908	Incorrect, Inadequate or Imprecise Result or Readings	10/07/2024	10/07/2024
							F08	Hospitalization or Prolonged Hospitalization	A0705	Battery Problem	10/07/2024	10/07/2024
									A070504	Premature Discharge of Battery	10/07/2024	10/07/2024
									A0908	Incorrect, Inadequate or Imprecise Result or Readings	10/07/2024	10/07/2024
							F1905	Device Revision or Replacement	A0705	Battery Problem	10/07/2024	10/07/2024
									A070504	Premature Discharge of Battery	10/07/2024	10/07/2024
									A0908	Incorrect, Inadequate or Imprecise Result or Readings	10/07/2024	10/07/2024
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem	10/07/2024	10/07/2024
									A070504	Premature Discharge of Battery	10/07/2024	10/07/2024
									A0908	Incorrect, Inadequate or Imprecise Result or Readings	10/07/2024	10/07/2024
							F08	Hospitalization or Prolonged Hospitalization	A0705	Battery Problem	10/07/2024	10/07/2024
									A070504	Premature Discharge of Battery	10/07/2024	10/07/2024
									A0908	Incorrect, Inadequate or Imprecise Result or Readings	10/07/2024	10/07/2024
							F1905	Device Revision or Replacement	A0705	Battery Problem	10/07/2024	10/07/2024
									A070504	Premature Discharge of Battery	10/07/2024	10/07/2024
									A0908	Incorrect, Inadequate or Imprecise Result or Readings	10/07/2024	10/07/2024
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem	10/07/2024	10/07/2024
									A070504	Premature Discharge of Battery	10/07/2024	10/07/2024
									A0908	Incorrect, Inadequate or Imprecise Result or Readings	10/07/2024	10/07/2024
							F08	Hospitalization or Prolonged Hospitalization	A0705	Battery Problem	10/07/2024	10/07/2024
									A070504	Premature Discharge of Battery	10/07/2024	10/07/2024
									A0908	Incorrect, Inadequate or Imprecise Result or Readings	10/07/2024	10/07/2024
							F1905	Device Revision or Replacement	A0705	Battery Problem	10/07/2024	10/07/2024
									A070504	Premature Discharge of Battery	10/07/2024	10/07/2024
									A0908	Incorrect, Inadequate or Imprecise Result or Readings	10/07/2024	10/07/2024
15526	2024	All other reportable incidents	47263	Cardiac resynchronization therapy implantable pacemaker	CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A072202	Low Impedance		
							F23	Unexpected Medical Intervention	A072202	Low Impedance		
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A072202	Low Impedance		
							F23	Unexpected Medical Intervention	A072202	Low Impedance		
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A072202	Low Impedance		
							F23	Unexpected Medical Intervention	A072202	Low Impedance		
15527	2024		37265	Dual-chamber implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A25	No Apparent Adverse Event	03/07/2024	03/07/2024
							F26	No Health Consequences or Impact	A25	No Apparent Adverse Event	03/07/2024	03/07/2024
15528	2024	All other reportable incidents	17882	Physiologic-monitoring defibrillation system	CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A0719	Unexpected Shutdown		
							F26	No Health Consequences or Impact	A0719	Unexpected Shutdown		
15529	2024	All other reportable incidents	47267	Single-chamber implantable pacemaker, rate-responsive	CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A071102	Failure to Interrogate		
							F26	No Health Consequences or Impact	A071102	Failure to Interrogate		
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A071102	Failure to Interrogate		
							F26	No Health Consequences or Impact	A071102	Failure to Interrogate		
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A071102	Failure to Interrogate		
							F26	No Health Consequences or Impact	A071102	Failure to Interrogate		
15530	2024	All other reportable incidents	47265	Dual-chamber implantable pacemaker, rate-responsive	CT1685	Cardiac pacemakers	E060101	Asystole	A0203	Defective Device		
							F24	Insufficient Health Impact Information	A0203	Defective Device		
					CT257	Implantable cardiac pacemakers	E060101	Asystole	A0203	Defective Device		
							F24	Insufficient Health Impact Information	A0203	Defective Device		
					CT2866	Implantable cardiac electrical stimulators	E060101	Asystole	A0203	Defective Device		
							F24	Insufficient Health Impact Information	A0203	Defective Device		
15531	2024	All other reportable incidents	17882	Physiologic-monitoring defibrillation system	CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070802	Intermittent Loss of Power		
							F27	Problem Identified During Non-Clinical Procedure	A070802	Intermittent Loss of Power		
15532	2024	All other reportable incidents	47267	Single-chamber implantable pacemaker, rate-responsive	CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A0908	Incorrect, Inadequate or Imprecise Result or Readings	13/08/2024	13/08/2024
							F08	Hospitalization or Prolonged Hospitalization	A0908	Incorrect, Inadequate or Imprecise Result or Readings	13/08/2024	13/08/2024
							F1903	Device Explantation	A0908	Incorrect, Inadequate or Imprecise Result or Readings	13/08/2024	13/08/2024
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A0908	Incorrect, Inadequate or Imprecise Result or Readings	13/08/2024	13/08/2024
							F08	Hospitalization or Prolonged Hospitalization	A0908	Incorrect, Inadequate or Imprecise Result or Readings	13/08/2024	13/08/2024
							F1903	Device Explantation	A0908	Incorrect, Inadequate or Imprecise Result or Readings	13/08/2024	13/08/2024
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A0908	Incorrect, Inadequate or Imprecise Result or Readings	13/08/2024	13/08/2024
							F08	Hospitalization or Prolonged Hospitalization	A0908	Incorrect, Inadequate or Imprecise Result or Readings	13/08/2024	13/08/2024
							F1903	Device Explantation	A0908	Incorrect, Inadequate or Imprecise Result or Readings	13/08/2024	13/08/2024
15533	2024	All other reportable incidents	47263	Cardiac resynchronization therapy implantable pacemaker	CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A0908	Incorrect, Inadequate or Imprecise Result or Readings	17/01/2024	17/01/2024
							F08	Hospitalization or Prolonged Hospitalization	A0908	Incorrect, Inadequate or Imprecise Result or Readings	17/01/2024	17/01/2024
							F1905	Device Revision or Replacement	A0908	Incorrect, Inadequate or Imprecise Result or Readings	17/01/2024	17/01/2024
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A0908	Incorrect, Inadequate or Imprecise Result or Readings	17/01/2024	17/01/2024

							F08	Hospitalization or Prolonged Hospitalization	A0908	Incorrect, Inadequate or Imprecise Result or Readings	17/01/2024	17/01/2024
							F1905	Device Revision or Replacement	A0908	Incorrect, Inadequate or Imprecise Result or Readings	17/01/2024	17/01/2024
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A0908	Incorrect, Inadequate or Imprecise Result or Readings	17/01/2024	17/01/2024
							F08	Hospitalization or Prolonged Hospitalization	A0908	Incorrect, Inadequate or Imprecise Result or Readings	17/01/2024	17/01/2024
							F1905	Device Revision or Replacement	A0908	Incorrect, Inadequate or Imprecise Result or Readings	17/01/2024	17/01/2024
15534	2024	All other reportable incidents	37265	Dual-chamber implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A072201	High Impedance		
									A0908	Incorrect, Inadequate or Imprecise Result or Readings		
							F26	No Health Consequences or Impact	A072201	High Impedance		
									A0908	Incorrect, Inadequate or Imprecise Result or Readings		
15535	2024		35822	External pacemaker, transcutaneous pacing	CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A0202	Defective Component		
							F26	No Health Consequences or Impact	A0202	Defective Component		
15536	2024		35822	External pacemaker, transcutaneous pacing	CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A0202	Defective Component		
							F26	No Health Consequences or Impact	A0202	Defective Component		
15537	2024		35822	External pacemaker, transcutaneous pacing	CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A0202	Defective Component		
									A0705	Battery Problem		
							F26	No Health Consequences or Impact	A0202	Defective Component		
									A0705	Battery Problem		
15538	2024	All other reportable incidents	47205	Cardiac pulse generator programmer	CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A0902	Display or Visual Feedback Problem		
									A23	Use of Device Problem		
							F26	No Health Consequences or Impact	A0902	Display or Visual Feedback Problem		
									A23	Use of Device Problem		
15539	2024	All other reportable incidents	47265	Dual-chamber implantable pacemaker, rate-responsive	CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem	01/03/2024	01/03/2024
									A070504	Premature Discharge of Battery	01/03/2024	01/03/2024
							F08	Hospitalization or Prolonged Hospitalization	A0705	Battery Problem	01/03/2024	01/03/2024
									A070504	Premature Discharge of Battery	01/03/2024	01/03/2024
							F1905	Device Revision or Replacement	A0705	Battery Problem	01/03/2024	01/03/2024
									A070504	Premature Discharge of Battery	01/03/2024	01/03/2024
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem	01/03/2024	01/03/2024
									A070504	Premature Discharge of Battery	01/03/2024	01/03/2024
							F08	Hospitalization or Prolonged Hospitalization	A0705	Battery Problem	01/03/2024	01/03/2024
									A070504	Premature Discharge of Battery	01/03/2024	01/03/2024
							F1905	Device Revision or Replacement	A0705	Battery Problem	01/03/2024	01/03/2024
									A070504	Premature Discharge of Battery	01/03/2024	01/03/2024
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem	01/03/2024	01/03/2024
									A070504	Premature Discharge of Battery	01/03/2024	01/03/2024
							F08	Hospitalization or Prolonged Hospitalization	A0705	Battery Problem	01/03/2024	01/03/2024
									A070504	Premature Discharge of Battery	01/03/2024	01/03/2024
							F1905	Device Revision or Replacement	A0705	Battery Problem	01/03/2024	01/03/2024
									A070504	Premature Discharge of Battery	01/03/2024	01/03/2024
15540	2024	All other reportable incidents	47265	Dual-chamber implantable pacemaker, rate-responsive	CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
15541	2024	All other reportable incidents	47270	Cardiac resynchronization therapy implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A070101	Failure to Capture		
							F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A070101	Failure to Capture		
15542	2024	All other reportable incidents	47265	Dual-chamber implantable pacemaker, rate-responsive	CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A070909	Oversensing		
									A0712	Pacing Problem		
									A090801	Signal Artifact/Noise		
							F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A070909	Oversensing		
									A0712	Pacing Problem		
									A090801	Signal Artifact/Noise		
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A070909	Oversensing		
									A0712	Pacing Problem		
									A090801	Signal Artifact/Noise		
							F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A070909	Oversensing		
									A0712	Pacing Problem		
									A090801	Signal Artifact/Noise		
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A070909	Oversensing		
									A0712	Pacing Problem		
									A090801	Signal Artifact/Noise		
							F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A070909	Oversensing		
									A0712	Pacing Problem		
									A090801	Signal Artifact/Noise		
15543	2024	All other reportable incidents	47263	Cardiac resynchronization therapy implantable pacemaker	CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A0908	Incorrect, Inadequate or Imprecise Result or Readings	05/07/2024	05/07/2024
							F08	Hospitalization or Prolonged Hospitalization	A0908	Incorrect, Inadequate or Imprecise Result or Readings	05/07/2024	05/07/2024
							F1905	Device Revision or Replacement	A0908	Incorrect, Inadequate or Imprecise Result or Readings	05/07/2024	05/07/2024
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A0908	Incorrect, Inadequate or Imprecise Result or Readings	05/07/2024	05/07/2024
							F08	Hospitalization or Prolonged Hospitalization	A0908	Incorrect, Inadequate or Imprecise Result or Readings	05/07/2024	05/07/2024
							F1905	Device Revision or Replacement	A0908	Incorrect, Inadequate or Imprecise Result or Readings	05/07/2024	05/07/2024

					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A0908	Incorrect, Inadequate or Imprecise Result or Readings	05/07/2024	05/07/2024
							F08	Hospitalization or Prolonged Hospitalization	A0908	Incorrect, Inadequate or Imprecise Result or Readings	05/07/2024	05/07/2024
							F1905	Device Revision or Replacement	A0908	Incorrect, Inadequate or Imprecise Result or Readings	05/07/2024	05/07/2024
15544	2024	All other reportable incidents	46842	Cardiac pulse generator analyser cable, single-use	CT1684	Cardiac pacemakers and associated devices	E210401	Shock from Patient Lead(s)	A070909	Oversensing		
									A071302	Inappropriate/Inadequate Shock/Stimulation		
							F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A070909	Oversensing		
									A071302	Inappropriate/Inadequate Shock/Stimulation		
15545	2024	All other reportable incidents	47265	Dual-chamber implantable pacemaker, rate-responsive	CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A0712	Pacing Problem	10/07/2024	10/07/2024
									A0908	Incorrect, Inadequate or Imprecise Result or Readings	10/07/2024	10/07/2024
							F08	Hospitalization or Prolonged Hospitalization	A0712	Pacing Problem	10/07/2024	10/07/2024
									A0908	Incorrect, Inadequate or Imprecise Result or Readings	10/07/2024	10/07/2024
							F1001	Absence of Treatment	A0712	Pacing Problem	10/07/2024	10/07/2024
									A0908	Incorrect, Inadequate or Imprecise Result or Readings	10/07/2024	10/07/2024
							F1905	Device Revision or Replacement	A0712	Pacing Problem	10/07/2024	10/07/2024
									A0908	Incorrect, Inadequate or Imprecise Result or Readings	10/07/2024	10/07/2024
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A0712	Pacing Problem	10/07/2024	10/07/2024
									A0908	Incorrect, Inadequate or Imprecise Result or Readings	10/07/2024	10/07/2024
							F08	Hospitalization or Prolonged Hospitalization	A0712	Pacing Problem	10/07/2024	10/07/2024
									A0908	Incorrect, Inadequate or Imprecise Result or Readings	10/07/2024	10/07/2024
							F1001	Absence of Treatment	A0712	Pacing Problem	10/07/2024	10/07/2024
									A0908	Incorrect, Inadequate or Imprecise Result or Readings	10/07/2024	10/07/2024
							F1905	Device Revision or Replacement	A0712	Pacing Problem	10/07/2024	10/07/2024
									A0908	Incorrect, Inadequate or Imprecise Result or Readings	10/07/2024	10/07/2024
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A0712	Pacing Problem	10/07/2024	10/07/2024
									A0908	Incorrect, Inadequate or Imprecise Result or Readings	10/07/2024	10/07/2024
							F08	Hospitalization or Prolonged Hospitalization	A0712	Pacing Problem	10/07/2024	10/07/2024
									A0908	Incorrect, Inadequate or Imprecise Result or Readings	10/07/2024	10/07/2024
							F1001	Absence of Treatment	A0712	Pacing Problem	10/07/2024	10/07/2024
									A0908	Incorrect, Inadequate or Imprecise Result or Readings	10/07/2024	10/07/2024
							F1905	Device Revision or Replacement	A0712	Pacing Problem	10/07/2024	10/07/2024
									A0908	Incorrect, Inadequate or Imprecise Result or Readings	10/07/2024	10/07/2024
15546	2024	All other reportable incidents	47270	Cardiac resynchronization therapy implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E011902	Presyncope	A071301	Failure to Deliver Shock/Stimulation		
							E060109	Tachycardia	A071301	Failure to Deliver Shock/Stimulation		
							F05	Delay to Treatment/ Therapy	A071301	Failure to Deliver Shock/Stimulation		
							F1203	Life Threatening Illness or Injury	A071301	Failure to Deliver Shock/Stimulation		
15547	2024	All other reportable incidents	17882	Physiologic-monitoring defibrillation system	CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070802	Intermittent Loss of Power		
							F26	No Health Consequences or Impact	A070802	Intermittent Loss of Power		
15548	2024		35822	External pacemaker, transcutaneous pacing	CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A0202	Defective Component		
									A0705	Battery Problem		
							F26	No Health Consequences or Impact	A0202	Defective Component		
									A0705	Battery Problem		
15549	2024	All other reportable incidents	47265	Dual-chamber implantable pacemaker, rate-responsive	CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem	08/08/2024	08/08/2024
									A070504	Premature Discharge of Battery	08/08/2024	08/08/2024
							F08	Hospitalization or Prolonged Hospitalization	A0705	Battery Problem	08/08/2024	08/08/2024
									A070504	Premature Discharge of Battery	08/08/2024	08/08/2024
							F1905	Device Revision or Replacement	A0705	Battery Problem	08/08/2024	08/08/2024
									A070504	Premature Discharge of Battery	08/08/2024	08/08/2024
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem	08/08/2024	08/08/2024
									A070504	Premature Discharge of Battery	08/08/2024	08/08/2024
							F08	Hospitalization or Prolonged Hospitalization	A0705	Battery Problem	08/08/2024	08/08/2024
									A070504	Premature Discharge of Battery	08/08/2024	08/08/2024
							F1905	Device Revision or Replacement	A0705	Battery Problem	08/08/2024	08/08/2024
									A070504	Premature Discharge of Battery	08/08/2024	08/08/2024
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem	08/08/2024	08/08/2024
									A070504	Premature Discharge of Battery	08/08/2024	08/08/2024
							F08	Hospitalization or Prolonged Hospitalization	A0705	Battery Problem	08/08/2024	08/08/2024
									A070504	Premature Discharge of Battery	08/08/2024	08/08/2024
							F1905	Device Revision or Replacement	A0705	Battery Problem	08/08/2024	08/08/2024
									A070504	Premature Discharge of Battery	08/08/2024	08/08/2024
15550	2024	All other reportable incidents	47265	Dual-chamber implantable pacemaker, rate-responsive	CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
15551	2024	All other reportable incidents	47263	Cardiac resynchronization therapy implantable pacemaker	CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A072202	Low Impedance		
									A0908	Incorrect, Inadequate or Imprecise Result or Readings		
							F26	No Health Consequences or Impact	A072202	Low Impedance		
									A0908	Incorrect, Inadequate or Imprecise Result or Readings		
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A072202	Low Impedance		
									A0908	Incorrect, Inadequate or Imprecise Result or Readings		
							F26	No Health Consequences or Impact	A072202	Low Impedance		
									A0908	Incorrect, Inadequate or Imprecise Result or Readings		

					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A072202	Low Impedance			
									A0908	Incorrect, Inadequate or Imprecise Result or Readings			
							F26	No Health Consequences or Impact	A072202	Low Impedance			
									A0908	Incorrect, Inadequate or Imprecise Result or Readings			
15552	2024	All other reportable incidents	47205	Cardiac pulse generator programmer	CT1684	Cardiac pacemakers and associated devices	E060101	Asystole	A0712	Pacing Problem			
									A0902	Display or Visual Feedback Problem			
									A13	Communication or Transmission Problem			
							E060104	Bradycardia	A0712	Pacing Problem			
									A0902	Display or Visual Feedback Problem			
									A13	Communication or Transmission Problem			
							F1203	Life Threatening Illness or Injury	A0712	Pacing Problem			
									A0902	Display or Visual Feedback Problem			
									A13	Communication or Transmission Problem			
15553	2024	All other reportable incidents	36106	Implantable pacemaker analyser, intraoperative	CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A1104	Computer Operating System Problem			
							F27	Problem Identified During Non-Clinical Procedure	A1104	Computer Operating System Problem			
15554	2024	All other reportable incidents	47270	Cardiac resynchronization therapy implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A0508	Noise, Audible	19/09/2024	19/09/2024	
									A070504	Premature Discharge of Battery	19/09/2024	19/09/2024	
									A0908	Incorrect, Inadequate or Imprecise Result or Readings	19/09/2024	19/09/2024	
							F08	Hospitalization or Prolonged Hospitalization	A0508	Noise, Audible	19/09/2024	19/09/2024	
									A070504	Premature Discharge of Battery	19/09/2024	19/09/2024	
									A0908	Incorrect, Inadequate or Imprecise Result or Readings	19/09/2024	19/09/2024	
							F1905	Device Revision or Replacement	A0508	Noise, Audible	19/09/2024	19/09/2024	
									A070504	Premature Discharge of Battery	19/09/2024	19/09/2024	
									A0908	Incorrect, Inadequate or Imprecise Result or Readings	19/09/2024	19/09/2024	
15555	2024	All other reportable incidents	36106	Implantable pacemaker analyser, intraoperative	CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A11	Computer Software Problem			
							F26	No Health Consequences or Impact	A11	Computer Software Problem			
15556	2024	All other reportable incidents	36106	Implantable pacemaker analyser, intraoperative	CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A11	Computer Software Problem			
							F26	No Health Consequences or Impact	A11	Computer Software Problem			
15557	2024	All other reportable incidents	61499	Vascular catheter introduction kit, steerable, single-use	CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A040607	Material Too Rigid or Stiff			
							F26	No Health Consequences or Impact	A040607	Material Too Rigid or Stiff			
15558	2024	All other reportable incidents	37265	Dual-chamber implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E0601	Arrhythmia	A071301	Failure to Deliver Shock/Stimulation			
									A090805	Incorrect Measurement			
							F23	Unexpected Medical Intervention	A071301	Failure to Deliver Shock/Stimulation			
									A090805	Incorrect Measurement			
15559	2024	All other reportable incidents	37265	Dual-chamber implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E060104	Bradycardia	A0712	Pacing Problem			
									A072001	Electromagnetic Interference			
							F11	Non-Serious Injury/ Illness/ Impairment	A0712	Pacing Problem			
									A072001	Electromagnetic Interference			
15560	2024	All other reportable incidents	47265	Dual-chamber implantable pacemaker, rate-responsive	CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	01/02/2024	01/02/2024	
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	01/02/2024	01/02/2024	
							F1903	Device Explantation	A070504	Premature Discharge of Battery	01/02/2024	01/02/2024	
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	01/02/2024	01/02/2024	
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	01/02/2024	01/02/2024	
							F1903	Device Explantation	A070504	Premature Discharge of Battery	01/02/2024	01/02/2024	
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	01/02/2024	01/02/2024	
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	01/02/2024	01/02/2024	
							F1903	Device Explantation	A070504	Premature Discharge of Battery	01/02/2024	01/02/2024	
15561	2024		45966	Transthoracic pacing lead	CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A0401	Break			
									A0501	Detachment of Device or Device Component			
							F2301	Additional Device Required	A0401	Break			
									A0501	Detachment of Device or Device Component			
							F26	No Health Consequences or Impact	A0401	Break			
									A0501	Detachment of Device or Device Component			
15562	2024		47265	Dual-chamber implantable pacemaker, rate-responsive	CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A0203	Defective Device	16/08/2024	16/08/2024	
									A070909	Oversensing	16/08/2024	16/08/2024	
							F1901	Additional Surgery	A0203	Defective Device	16/08/2024	16/08/2024	
									A070909	Oversensing	16/08/2024	16/08/2024	
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A0203	Defective Device	16/08/2024	16/08/2024	
									A070909	Oversensing	16/08/2024	16/08/2024	
							F1901	Additional Surgery	A0203	Defective Device	16/08/2024	16/08/2024	
									A070909	Oversensing	16/08/2024	16/08/2024	
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A0203	Defective Device	16/08/2024	16/08/2024	
									A070909	Oversensing	16/08/2024	16/08/2024	
							F1901	Additional Surgery	A0203	Defective Device	16/08/2024	16/08/2024	
									A070909	Oversensing	16/08/2024	16/08/2024	
15563	2024	All other reportable incidents	37265	Dual-chamber implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	29/03/2023	29/03/2023	
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	29/03/2023	29/03/2023	
							F12	Serious Injury/ Illness/ Impairment	A070504	Premature Discharge of Battery	29/03/2023	29/03/2023	
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	29/03/2023	29/03/2023	
15564	2024	All other reportable incidents	17882	Physiologic-monitoring defibrillation system	CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A071301	Failure to Deliver Shock/Stimulation			
							F27	Problem Identified During Non-Clinical Procedure	A071301	Failure to Deliver Shock/Stimulation			
15565	2024	All other reportable incidents	47265	Dual-chamber implantable pacemaker, rate-responsive	CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	05/09/2024	05/09/2024	
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	05/09/2024	05/09/2024	
							F12	Serious Injury/ Illness/ Impairment	A070504	Premature Discharge of Battery	05/09/2024	05/09/2024	
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	05/09/2024	05/09/2024	
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery	05/09/2024	05/09/2024	
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	05/09/2024	05/09/2024	
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	05/09/2024	05/09/2024	
							F12	Serious Injury/ Illness/ Impairment	A070504	Premature Discharge of Battery	05/09/2024	05/09/2024	
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	05/09/2024	05/09/2024	
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery	05/09/2024	05/09/2024	
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	05/09/2024	05/09/2024	
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	05/09/2024	05/09/2024	
							F12	Serious Injury/ Illness/ Impairment	A070504	Premature Discharge of Battery	05/09/2024	05/09/2024	

							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	05/09/2024	05/09/2024
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery	05/09/2024	05/09/2024
15566	2024	All other reportable incidents	35224	External pacemaker, epicardial pacing	CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A0701	Capturing Problem		
									A070910	Undersensing		
							F26	No Health Consequences or Impact	A0701	Capturing Problem		
									A070910	Undersensing		
15567	2024	All other reportable incidents	47267	Single-chamber implantable pacemaker, rate-responsive	CT1685	Cardiac pacemakers	E060109	Tachycardia	A0712	Pacing Problem		
									A16	Protective Measures Problem		
							F08	Hospitalization or Prolonged Hospitalization	A0712	Pacing Problem		
									A16	Protective Measures Problem		
							F12	Serious Injury/ Illness/ Impairment	A0712	Pacing Problem		
									A16	Protective Measures Problem		
							F2307	Reprogramming of Device	A0712	Pacing Problem		
									A16	Protective Measures Problem		
					CT257	Implantable cardiac pacemakers	E060109	Tachycardia	A0712	Pacing Problem		
									A16	Protective Measures Problem		
							F08	Hospitalization or Prolonged Hospitalization	A0712	Pacing Problem		
									A16	Protective Measures Problem		
							F12	Serious Injury/ Illness/ Impairment	A0712	Pacing Problem		
									A16	Protective Measures Problem		
							F2307	Reprogramming of Device	A0712	Pacing Problem		
									A16	Protective Measures Problem		
					CT2866	Implantable cardiac electrical stimulators	E060109	Tachycardia	A0712	Pacing Problem		
									A16	Protective Measures Problem		
							F08	Hospitalization or Prolonged Hospitalization	A0712	Pacing Problem		
									A16	Protective Measures Problem		
							F12	Serious Injury/ Illness/ Impairment	A0712	Pacing Problem		
									A16	Protective Measures Problem		
							F2307	Reprogramming of Device	A0712	Pacing Problem		
									A16	Protective Measures Problem		
15568	2024	All other reportable incidents	35852	Single-chamber implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A070909	Oversensing		
									A090801	Signal Artifact/Noise		
									A1301	Failure to Read Input Signal		
							F2203	Imaging Required	A070909	Oversensing		
									A090801	Signal Artifact/Noise		
									A1301	Failure to Read Input Signal		
15569	2024	All other reportable incidents	47270	Cardiac resynchronization therapy implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E210401	Shock from Patient Lead(s)	A0203	Defective Device	18/07/2024	18/07/2024
									A040101	Fracture	18/07/2024	18/07/2024
									A070909	Oversensing	18/07/2024	18/07/2024
									A071302	Inappropriate/Inadequate Shock/Stimulation	18/07/2024	18/07/2024
									A090801	Signal Artifact/Noise	18/07/2024	18/07/2024
							F08	Hospitalization or Prolonged Hospitalization	A0203	Defective Device	18/07/2024	18/07/2024
									A040101	Fracture	18/07/2024	18/07/2024
									A070909	Oversensing	18/07/2024	18/07/2024
									A071302	Inappropriate/Inadequate Shock/Stimulation	18/07/2024	18/07/2024
									A090801	Signal Artifact/Noise	18/07/2024	18/07/2024
							F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A0203	Defective Device	18/07/2024	18/07/2024
									A040101	Fracture	18/07/2024	18/07/2024
									A070909	Oversensing	18/07/2024	18/07/2024
									A071302	Inappropriate/Inadequate Shock/Stimulation	18/07/2024	18/07/2024
									A090801	Signal Artifact/Noise	18/07/2024	18/07/2024
							F1905	Device Revision or Replacement	A0203	Defective Device	18/07/2024	18/07/2024
									A040101	Fracture	18/07/2024	18/07/2024
									A070909	Oversensing	18/07/2024	18/07/2024
									A071302	Inappropriate/Inadequate Shock/Stimulation	18/07/2024	18/07/2024
									A090801	Signal Artifact/Noise	18/07/2024	18/07/2024
15570	2024	All other reportable incidents	47265	Dual-chamber implantable pacemaker, rate-responsive	CT1685	Cardiac pacemakers	E0623	Unspecified Heart Problem	A071209	Extracardiac Stimulation	19/01/2024	19/01/2024
							E2103	Device Overstimulation of Tissue	A071209	Extracardiac Stimulation	19/01/2024	19/01/2024
							F08	Hospitalization or Prolonged Hospitalization	A071209	Extracardiac Stimulation	19/01/2024	19/01/2024
							F1905	Device Revision or Replacement	A071209	Extracardiac Stimulation	19/01/2024	19/01/2024
					CT257	Implantable cardiac pacemakers	E0623	Unspecified Heart Problem	A071209	Extracardiac Stimulation	19/01/2024	19/01/2024
							E2103	Device Overstimulation of Tissue	A071209	Extracardiac Stimulation	19/01/2024	19/01/2024
							F08	Hospitalization or Prolonged Hospitalization	A071209	Extracardiac Stimulation	19/01/2024	19/01/2024
							F1905	Device Revision or Replacement	A071209	Extracardiac Stimulation	19/01/2024	19/01/2024
					CT2866	Implantable cardiac electrical stimulators	E0623	Unspecified Heart Problem	A071209	Extracardiac Stimulation	19/01/2024	19/01/2024
							E2103	Device Overstimulation of Tissue	A071209	Extracardiac Stimulation	19/01/2024	19/01/2024
							F08	Hospitalization or Prolonged Hospitalization	A071209	Extracardiac Stimulation	19/01/2024	19/01/2024
							F1905	Device Revision or Replacement	A071209	Extracardiac Stimulation	19/01/2024	19/01/2024
15571	2024	All other reportable incidents	47270	Cardiac resynchronization therapy implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem	16/07/2024	16/07/2024
									A070504	Premature Discharge of Battery	16/07/2024	16/07/2024
									A071102	Failure to Interrogate	16/07/2024	16/07/2024
							F26	No Health Consequences or Impact	A0705	Battery Problem	16/07/2024	16/07/2024
									A070504	Premature Discharge of Battery	16/07/2024	16/07/2024
									A071102	Failure to Interrogate	16/07/2024	16/07/2024
15572	2024	All other reportable incidents	47265	Dual-chamber implantable pacemaker, rate-responsive	CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
15573	2024	All other reportable incidents	47205	Cardiac pulse generator programmer	CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A0203	Defective Device		
									A070803	Failure to Power Up		

									A150105	Key or Button Unresponsive/not Working		
							F26	No Health Consequences or Impact	A0203	Defective Device		
									A070803	Failure to Power Up		
									A150105	Key or Button Unresponsive/not Working		
15574	2024	All other reportable incidents	35852	Single-chamber implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
									A0908	Incorrect, Inadequate or Imprecise Result or Readings		
							F26	No Health Consequences or Impact	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
									A0908	Incorrect, Inadequate or Imprecise Result or Readings		
15575	2024	All other reportable incidents	47265	Dual-chamber implantable pacemaker, rate-responsive	CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
15576	2024	All other reportable incidents	47265	Dual-chamber implantable pacemaker, rate-responsive	CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	22/07/2024	22/07/2024
							F1901	Additional Surgery	A24	No Apparent Device Problem	22/07/2024	22/07/2024
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	22/07/2024	22/07/2024
							F1901	Additional Surgery	A24	No Apparent Device Problem	22/07/2024	22/07/2024
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	22/07/2024	22/07/2024
							F1901	Additional Surgery	A24	No Apparent Device Problem	22/07/2024	22/07/2024
15577	2024	All other reportable incidents	47267	Single-chamber implantable pacemaker, rate-responsive	CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A070909	Oversensing	19/07/2024	19/07/2024
									A0712	Pacing Problem	19/07/2024	19/07/2024
									A090801	Signal Artifact/Noise	19/07/2024	19/07/2024
							F26	No Health Consequences or Impact	A070909	Oversensing	19/07/2024	19/07/2024
									A0712	Pacing Problem	19/07/2024	19/07/2024
									A090801	Signal Artifact/Noise	19/07/2024	19/07/2024
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A070909	Oversensing	19/07/2024	19/07/2024
									A0712	Pacing Problem	19/07/2024	19/07/2024
									A090801	Signal Artifact/Noise	19/07/2024	19/07/2024
							F26	No Health Consequences or Impact	A070909	Oversensing	19/07/2024	19/07/2024
									A0712	Pacing Problem	19/07/2024	19/07/2024
									A090801	Signal Artifact/Noise	19/07/2024	19/07/2024
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A070909	Oversensing	19/07/2024	19/07/2024
									A0712	Pacing Problem	19/07/2024	19/07/2024
									A090801	Signal Artifact/Noise	19/07/2024	19/07/2024
							F26	No Health Consequences or Impact	A070909	Oversensing	19/07/2024	19/07/2024
									A0712	Pacing Problem	19/07/2024	19/07/2024
									A090801	Signal Artifact/Noise	19/07/2024	19/07/2024
15578	2024	All other reportable incidents	47270	Cardiac resynchronization therapy implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A070909	Oversensing		
									A0908	Incorrect, Inadequate or Imprecise Result or Readings		
							F26	No Health Consequences or Impact	A070909	Oversensing		
									A0908	Incorrect, Inadequate or Imprecise Result or Readings		
15579	2024	All other reportable incidents	47265	Dual-chamber implantable pacemaker, rate-responsive	CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
15580	2024	All other reportable incidents	47265	Dual-chamber implantable pacemaker, rate-responsive	CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	12/02/2024	12/02/2024
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	12/02/2024	12/02/2024
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	12/02/2024	12/02/2024
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	12/02/2024	12/02/2024
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	12/02/2024	12/02/2024
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	12/02/2024	12/02/2024
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	12/02/2024	12/02/2024
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	12/02/2024	12/02/2024
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	12/02/2024	12/02/2024
15581	2024	All other reportable incidents	37265	Dual-chamber implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	11/11/2024	11/11/2024
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	11/11/2024	11/11/2024
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	11/11/2024	11/11/2024
15582	2024	All other reportable incidents	45966	Transthoracic pacing lead	CT1684	Cardiac pacemakers and associated devices	E0601	Arrhythmia	A070910	Undersensing		
							E0602	Cardiac Arrest	A070910	Undersensing		
							F23	Unexpected Medical Intervention	A070910	Undersensing		
15583	2024	All other reportable incidents	17882	Physiologic-monitoring defibrillation system	CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A071301	Failure to Deliver Shock/Stimulation		
							F27	Problem Identified During Non-Clinical Procedure	A071301	Failure to Deliver Shock/Stimulation		
15584	2024	All other reportable incidents	47265	Dual-chamber implantable pacemaker, rate-responsive	CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A0908	Incorrect, Inadequate or Imprecise Result or Readings	18/07/2024	18/07/2024
							F08	Hospitalization or Prolonged Hospitalization	A0908	Incorrect, Inadequate or Imprecise Result or Readings	18/07/2024	18/07/2024
							F1905	Device Revision or Replacement	A0908	Incorrect, Inadequate or Imprecise Result or Readings	18/07/2024	18/07/2024
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A0908	Incorrect, Inadequate or Imprecise Result or Readings	18/07/2024	18/07/2024
							F08	Hospitalization or Prolonged Hospitalization	A0908	Incorrect, Inadequate or Imprecise Result or Readings	18/07/2024	18/07/2024
							F1905	Device Revision or Replacement	A0908	Incorrect, Inadequate or Imprecise Result or Readings	18/07/2024	18/07/2024

					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A0908	Incorrect, Inadequate or Imprecise Result or Readings	18/07/2024	18/07/2024
							F08	Hospitalization or Prolonged Hospitalization	A0908	Incorrect, Inadequate or Imprecise Result or Readings	18/07/2024	18/07/2024
							F1905	Device Revision or Replacement	A0908	Incorrect, Inadequate or Imprecise Result or Readings	18/07/2024	18/07/2024
15585	2024	All other reportable incidents	47270	Cardiac resynchronization therapy implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	27/06/2023	27/06/2023
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	27/06/2023	27/06/2023
							F12	Serious Injury/ Illness/ Impairment	A070504	Premature Discharge of Battery	27/06/2023	27/06/2023
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	27/06/2023	27/06/2023
15586	2024	All other reportable incidents	17882	Physiologic-monitoring defibrillation system	CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A11	Computer Software Problem		
							F27	Problem Identified During Non-Clinical Procedure	A11	Computer Software Problem		
15587	2024	All other reportable incidents	47205	Cardiac pulse generator programmer	CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A09	Output Problem		
									A13	Communication or Transmission Problem		
							F26	No Health Consequences or Impact	A09	Output Problem		
									A13	Communication or Transmission Problem		
15588	2024	All other reportable incidents	47205	Cardiac pulse generator programmer	CT1684	Cardiac pacemakers and associated devices	E210401	Shock from Patient Lead(s)	A1104	Computer Operating System Problem		
							F23	Unexpected Medical Intervention	A1104	Computer Operating System Problem		
15589	2024	All other reportable incidents	47265	Dual-chamber implantable pacemaker, rate-responsive	CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
15590	2024	All other reportable incidents	37265	Dual-chamber implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	05/12/2024	05/12/2024
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	05/12/2024	05/12/2024
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	05/12/2024	05/12/2024
15591	2024	All other reportable incidents	47265	Dual-chamber implantable pacemaker, rate-responsive	CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem		
							F1901	Additional Surgery	A24	No Apparent Device Problem		
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem		
							F1901	Additional Surgery	A24	No Apparent Device Problem		
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem		
							F1901	Additional Surgery	A24	No Apparent Device Problem		
15592	2024	All other reportable incidents	47270	Cardiac resynchronization therapy implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E210401	Shock from Patient Lead(s)	A0709	Device Sensing Problem	20/09/2024	20/09/2024
									A071302	Inappropriate/Inadequate Shock/Stimulation	20/09/2024	20/09/2024
									A072201	High Impedance	20/09/2024	20/09/2024
									A0908	Incorrect, Inadequate or Imprecise Result or Readings	20/09/2024	20/09/2024
							F08	Hospitalization or Prolonged Hospitalization	A0709	Device Sensing Problem	20/09/2024	20/09/2024
									A071302	Inappropriate/Inadequate Shock/Stimulation	20/09/2024	20/09/2024
									A072201	High Impedance	20/09/2024	20/09/2024
									A0908	Incorrect, Inadequate or Imprecise Result or Readings	20/09/2024	20/09/2024
							F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A0709	Device Sensing Problem	20/09/2024	20/09/2024
									A071302	Inappropriate/Inadequate Shock/Stimulation	20/09/2024	20/09/2024
									A072201	High Impedance	20/09/2024	20/09/2024
									A0908	Incorrect, Inadequate or Imprecise Result or Readings	20/09/2024	20/09/2024
							F1905	Device Revision or Replacement	A0709	Device Sensing Problem	20/09/2024	20/09/2024
									A071302	Inappropriate/Inadequate Shock/Stimulation	20/09/2024	20/09/2024
									A072201	High Impedance	20/09/2024	20/09/2024
									A0908	Incorrect, Inadequate or Imprecise Result or Readings	20/09/2024	20/09/2024
							F22	Unexpected Diagnostic Intervention	A0709	Device Sensing Problem	20/09/2024	20/09/2024
									A071302	Inappropriate/Inadequate Shock/Stimulation	20/09/2024	20/09/2024
									A072201	High Impedance	20/09/2024	20/09/2024
									A0908	Incorrect, Inadequate or Imprecise Result or Readings	20/09/2024	20/09/2024
							F2203	Imaging Required	A0709	Device Sensing Problem	20/09/2024	20/09/2024
									A071302	Inappropriate/Inadequate Shock/Stimulation	20/09/2024	20/09/2024
									A072201	High Impedance	20/09/2024	20/09/2024
									A0908	Incorrect, Inadequate or Imprecise Result or Readings	20/09/2024	20/09/2024
15593	2024		37265	Dual-chamber implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A0201	Product Quality Problem		
									A0203	Defective Device		
							F1905	Device Revision or Replacement	A0201	Product Quality Problem		
									A0203	Defective Device		
15594	2024	All other reportable incidents			CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070909	Oversensing		
									A070910	Undersensing		
									A1301	Failure to Read Input Signal		
							F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A070909	Oversensing		
									A070910	Undersensing		
									A1301	Failure to Read Input Signal		
15595	2024	All other reportable incidents	17882	Physiologic-monitoring defibrillation system	CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A0908	Incorrect, Inadequate or Imprecise Result or Readings		
							F27	Problem Identified During Non-Clinical Procedure	A0908	Incorrect, Inadequate or Imprecise Result or Readings		
15596	2024	All other reportable incidents	17882	Physiologic-monitoring defibrillation system	CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A150105	Key or Button Unresponsive/not Working		
							F27	Problem Identified During Non-Clinical Procedure	A150105	Key or Button Unresponsive/not Working		
15597	2024	All other reportable incidents	47205	Cardiac pulse generator programmer	CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A0902	Display or Visual Feedback Problem		
									A11	Computer Software Problem		
									A13	Communication or Transmission Problem		
							F26	No Health Consequences or Impact	A0902	Display or Visual Feedback Problem		
									A11	Computer Software Problem		
									A13	Communication or Transmission Problem		
15598	2024		47265	Dual-chamber implantable pacemaker, rate-responsive	CT1685	Cardiac pacemakers	E2401	Insufficient Information	A0705	Battery Problem		
							F11	Non-Serious Injury/ Illness/ Impairment	A0705	Battery Problem		
					CT257	Implantable cardiac pacemakers	E2401	Insufficient Information	A0705	Battery Problem		
							F11	Non-Serious Injury/ Illness/ Impairment	A0705	Battery Problem		
					CT2866	Implantable cardiac electrical stimulators	E2401	Insufficient Information	A0705	Battery Problem		

							F11	Non-Serious Injury/ Illness/ Impairment	A0705	Battery Problem		
15599	2024	All other reportable incidents	17882	Physiologic-monitoring defibrillation system	CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070802	Intermittent Loss of Power		
							F27	Problem Identified During Non-Clinical Procedure	A070802	Intermittent Loss of Power		
15600	2024	All other reportable incidents	17882	Physiologic-monitoring defibrillation system	CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A07	Electrical /Electronic Property Problem		
							F27	Problem Identified During Non-Clinical Procedure	A07	Electrical /Electronic Property Problem		
15601	2024	All other reportable incidents	47270	Cardiac resynchronization therapy implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	24/05/2024	24/05/2024
							F1901	Additional Surgery	A070504	Premature Discharge of Battery	24/05/2024	24/05/2024
15602	2024	All other reportable incidents			CT1684	Cardiac pacemakers and associated devices	E060101	Asystole	A070909	Oversensing		
									A0712	Pacing Problem		
									A090801	Signal Artifact/Noise		
							F1001	Absence of Treatment	A070909	Oversensing		
									A0712	Pacing Problem		
									A090801	Signal Artifact/Noise		
15603	2024	All other reportable incidents	47263	Cardiac resynchronization therapy implantable pacemaker	CT1685	Cardiac pacemakers	E060101	Asystole	A070909	Oversensing	24/07/2024	24/07/2024
									A0712	Pacing Problem	24/07/2024	24/07/2024
									A0908	Incorrect, Inadequate or Imprecise Result or Readings	24/07/2024	24/07/2024
							F08	Hospitalization or Prolonged Hospitalization	A070909	Oversensing	24/07/2024	24/07/2024
									A0712	Pacing Problem	24/07/2024	24/07/2024
									A0908	Incorrect, Inadequate or Imprecise Result or Readings	24/07/2024	24/07/2024
							F1001	Absence of Treatment	A070909	Oversensing	24/07/2024	24/07/2024
									A0712	Pacing Problem	24/07/2024	24/07/2024
									A0908	Incorrect, Inadequate or Imprecise Result or Readings	24/07/2024	24/07/2024
							F1905	Device Revision or Replacement	A070909	Oversensing	24/07/2024	24/07/2024
									A0712	Pacing Problem	24/07/2024	24/07/2024
									A0908	Incorrect, Inadequate or Imprecise Result or Readings	24/07/2024	24/07/2024
					CT257	Implantable cardiac pacemakers	E060101	Asystole	A070909	Oversensing	24/07/2024	24/07/2024
									A0712	Pacing Problem	24/07/2024	24/07/2024
									A0908	Incorrect, Inadequate or Imprecise Result or Readings	24/07/2024	24/07/2024
							F08	Hospitalization or Prolonged Hospitalization	A070909	Oversensing	24/07/2024	24/07/2024
									A0712	Pacing Problem	24/07/2024	24/07/2024
									A0908	Incorrect, Inadequate or Imprecise Result or Readings	24/07/2024	24/07/2024
							F1001	Absence of Treatment	A070909	Oversensing	24/07/2024	24/07/2024
									A0712	Pacing Problem	24/07/2024	24/07/2024
									A0908	Incorrect, Inadequate or Imprecise Result or Readings	24/07/2024	24/07/2024
							F1905	Device Revision or Replacement	A070909	Oversensing	24/07/2024	24/07/2024
									A0712	Pacing Problem	24/07/2024	24/07/2024
									A0908	Incorrect, Inadequate or Imprecise Result or Readings	24/07/2024	24/07/2024
					CT2866	Implantable cardiac electrical stimulators	E060101	Asystole	A070909	Oversensing	24/07/2024	24/07/2024
									A0712	Pacing Problem	24/07/2024	24/07/2024
									A0908	Incorrect, Inadequate or Imprecise Result or Readings	24/07/2024	24/07/2024
							F08	Hospitalization or Prolonged Hospitalization	A070909	Oversensing	24/07/2024	24/07/2024
									A0712	Pacing Problem	24/07/2024	24/07/2024
									A0908	Incorrect, Inadequate or Imprecise Result or Readings	24/07/2024	24/07/2024
							F1001	Absence of Treatment	A070909	Oversensing	24/07/2024	24/07/2024
									A0712	Pacing Problem	24/07/2024	24/07/2024
									A0908	Incorrect, Inadequate or Imprecise Result or Readings	24/07/2024	24/07/2024
							F1905	Device Revision or Replacement	A070909	Oversensing	24/07/2024	24/07/2024
									A0712	Pacing Problem	24/07/2024	24/07/2024
									A0908	Incorrect, Inadequate or Imprecise Result or Readings	24/07/2024	24/07/2024
15604	2024	All other reportable incidents	47265	Dual-chamber implantable pacemaker, rate-responsive	CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
15605	2024	All other reportable incidents	60789	Intracardiac pacemaker	CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT2818	Catheterization/Cannulation/Fluid path devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
15606	2024	All other reportable incidents	47205	Cardiac pulse generator programmer	CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A11	Computer Software Problem		
									A2204	Inadequate User Interface		
							F26	No Health Consequences or Impact	A11	Computer Software Problem		
									A2204	Inadequate User Interface		
15607	2024	All other reportable incidents	35852	Single-chamber implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E210401	Shock from Patient Lead(s)	A070909	Oversensing		
									A071302	Inappropriate/Inadequate Shock/Stimulation		
									A1301	Failure to Read Input Signal		
							F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A070909	Oversensing		
									A071302	Inappropriate/Inadequate Shock/Stimulation		
									A1301	Failure to Read Input Signal		

15608	2024	All other reportable incidents	47265	Dual-chamber implantable pacemaker, rate-responsive	CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
15609	2024	All other reportable incidents	47263	Cardiac resynchronization therapy implantable pacemaker	CT1685	Cardiac pacemakers	E0112	Dizziness	A070101	Failure to Capture	13/04/2024	13/04/2024
									A0908	Incorrect, Inadequate or Imprecise Result or Readings	13/04/2024	13/04/2024
							F08	Hospitalization or Prolonged Hospitalization	A070101	Failure to Capture	13/04/2024	13/04/2024
									A0908	Incorrect, Inadequate or Imprecise Result or Readings	13/04/2024	13/04/2024
							F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A070101	Failure to Capture	13/04/2024	13/04/2024
									A0908	Incorrect, Inadequate or Imprecise Result or Readings	13/04/2024	13/04/2024
							F1905	Device Revision or Replacement	A070101	Failure to Capture	13/04/2024	13/04/2024
									A0908	Incorrect, Inadequate or Imprecise Result or Readings	13/04/2024	13/04/2024
					CT257	Implantable cardiac pacemakers	E0112	Dizziness	A070101	Failure to Capture	13/04/2024	13/04/2024
									A0908	Incorrect, Inadequate or Imprecise Result or Readings	13/04/2024	13/04/2024
							F08	Hospitalization or Prolonged Hospitalization	A070101	Failure to Capture	13/04/2024	13/04/2024
									A0908	Incorrect, Inadequate or Imprecise Result or Readings	13/04/2024	13/04/2024
							F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A070101	Failure to Capture	13/04/2024	13/04/2024
									A0908	Incorrect, Inadequate or Imprecise Result or Readings	13/04/2024	13/04/2024
							F1905	Device Revision or Replacement	A070101	Failure to Capture	13/04/2024	13/04/2024
									A0908	Incorrect, Inadequate or Imprecise Result or Readings	13/04/2024	13/04/2024
					CT2866	Implantable cardiac electrical stimulators	E0112	Dizziness	A070101	Failure to Capture	13/04/2024	13/04/2024
									A0908	Incorrect, Inadequate or Imprecise Result or Readings	13/04/2024	13/04/2024
							F08	Hospitalization or Prolonged Hospitalization	A070101	Failure to Capture	13/04/2024	13/04/2024
									A0908	Incorrect, Inadequate or Imprecise Result or Readings	13/04/2024	13/04/2024
							F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A070101	Failure to Capture	13/04/2024	13/04/2024
									A0908	Incorrect, Inadequate or Imprecise Result or Readings	13/04/2024	13/04/2024
							F1905	Device Revision or Replacement	A070101	Failure to Capture	13/04/2024	13/04/2024
									A0908	Incorrect, Inadequate or Imprecise Result or Readings	13/04/2024	13/04/2024
15610	2024	All other reportable incidents	47265	Dual-chamber implantable pacemaker, rate-responsive	CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A0908	Incorrect, Inadequate or Imprecise Result or Readings	01/07/2024	01/07/2024
							F08	Hospitalization or Prolonged Hospitalization	A0908	Incorrect, Inadequate or Imprecise Result or Readings	01/07/2024	01/07/2024
							F1905	Device Revision or Replacement	A0908	Incorrect, Inadequate or Imprecise Result or Readings	01/07/2024	01/07/2024
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A0908	Incorrect, Inadequate or Imprecise Result or Readings	01/07/2024	01/07/2024
							F08	Hospitalization or Prolonged Hospitalization	A0908	Incorrect, Inadequate or Imprecise Result or Readings	01/07/2024	01/07/2024
							F1905	Device Revision or Replacement	A0908	Incorrect, Inadequate or Imprecise Result or Readings	01/07/2024	01/07/2024
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A0908	Incorrect, Inadequate or Imprecise Result or Readings	01/07/2024	01/07/2024
							F08	Hospitalization or Prolonged Hospitalization	A0908	Incorrect, Inadequate or Imprecise Result or Readings	01/07/2024	01/07/2024
							F1905	Device Revision or Replacement	A0908	Incorrect, Inadequate or Imprecise Result or Readings	01/07/2024	01/07/2024
15611	2024	All other reportable incidents	46842	Cardiac pulse generator analyser cable, single-use	CT1684	Cardiac pacemakers and associated devices	E210401	Shock from Patient Lead(s)	A070909	Oversensing		
									A071302	Inappropriate/Inadequate Shock/Stimulation		
							F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A070909	Oversensing		
									A071302	Inappropriate/Inadequate Shock/Stimulation		
							F2203	Imaging Required	A070909	Oversensing		
									A071302	Inappropriate/Inadequate Shock/Stimulation		
							F23	Unexpected Medical Intervention	A070909	Oversensing		
									A071302	Inappropriate/Inadequate Shock/Stimulation		
15612	2024	All other reportable incidents	47265	Dual-chamber implantable pacemaker, rate-responsive	CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A0908	Incorrect, Inadequate or Imprecise Result or Readings	11/07/2024	11/07/2024
							F08	Hospitalization or Prolonged Hospitalization	A0908	Incorrect, Inadequate or Imprecise Result or Readings	11/07/2024	11/07/2024
							F1905	Device Revision or Replacement	A0908	Incorrect, Inadequate or Imprecise Result or Readings	11/07/2024	11/07/2024
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A0908	Incorrect, Inadequate or Imprecise Result or Readings	11/07/2024	11/07/2024
							F08	Hospitalization or Prolonged Hospitalization	A0908	Incorrect, Inadequate or Imprecise Result or Readings	11/07/2024	11/07/2024
							F1905	Device Revision or Replacement	A0908	Incorrect, Inadequate or Imprecise Result or Readings	11/07/2024	11/07/2024
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A0908	Incorrect, Inadequate or Imprecise Result or Readings	11/07/2024	11/07/2024
							F08	Hospitalization or Prolonged Hospitalization	A0908	Incorrect, Inadequate or Imprecise Result or Readings	11/07/2024	11/07/2024
							F1905	Device Revision or Replacement	A0908	Incorrect, Inadequate or Imprecise Result or Readings	11/07/2024	11/07/2024
15613	2024	All other reportable incidents	47265	Dual-chamber implantable pacemaker, rate-responsive	CT1685	Cardiac pacemakers	E0112	Dizziness	A071102	Failure to Interrogate	08/08/2024	08/08/2024
							F1901	Additional Surgery	A071102	Failure to Interrogate	08/08/2024	08/08/2024
					CT257	Implantable cardiac pacemakers	E0112	Dizziness	A071102	Failure to Interrogate	08/08/2024	08/08/2024

					CT2866	Implantable cardiac electrical stimulators	F1901	Additional Surgery	A071102	Failure to Interrogate	08/08/2024	08/08/2024
							E0112	Dizziness	A071102	Failure to Interrogate	08/08/2024	08/08/2024
							F1901	Additional Surgery	A071102	Failure to Interrogate	08/08/2024	08/08/2024
15614	2024	All other reportable incidents	47205	Cardiac pulse generator programmer	CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A13	Communication or Transmission Problem		
							F26	No Health Consequences or Impact	A13	Communication or Transmission Problem		
15615	2024	All other reportable incidents	47265	Dual-chamber implantable pacemaker, rate-responsive	CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	23/07/2024	23/07/2024
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	23/07/2024	23/07/2024
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	23/07/2024	23/07/2024
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	23/07/2024	23/07/2024
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	23/07/2024	23/07/2024
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	23/07/2024	23/07/2024
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	23/07/2024	23/07/2024
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	23/07/2024	23/07/2024
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	23/07/2024	23/07/2024
15616	2024	Unanticipated serious deterioration in state of health			CT1684	Cardiac pacemakers and associated devices	E0602	Cardiac Arrest	A24	No Apparent Device Problem	22/07/2024	
							F02	Death	A24	No Apparent Device Problem	22/07/2024	
					CT257	Implantable cardiac pacemakers	E0602	Cardiac Arrest	A24	No Apparent Device Problem	22/07/2024	
							F02	Death	A24	No Apparent Device Problem	22/07/2024	
15617	2024	All other reportable incidents	47265	Dual-chamber implantable pacemaker, rate-responsive	CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
15618	2024	All other reportable incidents	47267	Single-chamber implantable pacemaker, rate-responsive	CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A072201	High Impedance		
							F26	No Health Consequences or Impact	A072201	High Impedance		
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A072201	High Impedance		
							F26	No Health Consequences or Impact	A072201	High Impedance		
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A072201	High Impedance		
							F26	No Health Consequences or Impact	A072201	High Impedance		
15619	2024	All other reportable incidents	17882	Physiologic-monitoring defibrillation system	CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A150105	Key or Button Unresponsive/not Working		
							F27	Problem Identified During Non-Clinical Procedure	A150105	Key or Button Unresponsive/not Working		
15620	2024	All other reportable incidents	47267	Single-chamber implantable pacemaker, rate-responsive	CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A0908	Incorrect, Inadequate or Imprecise Result or Readings	26/07/2024	26/07/2024
							F08	Hospitalization or Prolonged Hospitalization	A0908	Incorrect, Inadequate or Imprecise Result or Readings	26/07/2024	26/07/2024
							F1905	Device Revision or Replacement	A0908	Incorrect, Inadequate or Imprecise Result or Readings	26/07/2024	26/07/2024
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A0908	Incorrect, Inadequate or Imprecise Result or Readings	26/07/2024	26/07/2024
							F08	Hospitalization or Prolonged Hospitalization	A0908	Incorrect, Inadequate or Imprecise Result or Readings	26/07/2024	26/07/2024
							F1905	Device Revision or Replacement	A0908	Incorrect, Inadequate or Imprecise Result or Readings	26/07/2024	26/07/2024
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A0908	Incorrect, Inadequate or Imprecise Result or Readings	26/07/2024	26/07/2024
							F08	Hospitalization or Prolonged Hospitalization	A0908	Incorrect, Inadequate or Imprecise Result or Readings	26/07/2024	26/07/2024
							F1905	Device Revision or Replacement	A0908	Incorrect, Inadequate or Imprecise Result or Readings	26/07/2024	26/07/2024
15621	2024	All other reportable incidents	17882	Physiologic-monitoring defibrillation system	CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A071301	Failure to Deliver Shock/Stimulation		
							F27	Problem Identified During Non-Clinical Procedure	A071301	Failure to Deliver Shock/Stimulation		
15622	2024	All other reportable incidents	47265	Dual-chamber implantable pacemaker, rate-responsive	CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A070909	Oversensing		
							F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A070909	Oversensing		
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A070909	Oversensing		
							F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A070909	Oversensing		
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A070909	Oversensing		
							F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A070909	Oversensing		
15623	2024	All other reportable incidents	47265	Dual-chamber implantable pacemaker, rate-responsive	CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A071102	Failure to Interrogate		
							F26	No Health Consequences or Impact	A071102	Failure to Interrogate		
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A071102	Failure to Interrogate		
							F26	No Health Consequences or Impact	A071102	Failure to Interrogate		
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A071102	Failure to Interrogate		
							F26	No Health Consequences or Impact	A071102	Failure to Interrogate		
15624	2024	All other reportable incidents	60789	Intracardiac pacemaker	CT1685	Cardiac pacemakers	E2401	Insufficient Information	A0701	Capturing Problem		
									A070101	Failure to Capture		
									A070102	High Capture Threshold		
									A070104	Unstable Capture Threshold		
							F08	Hospitalization or Prolonged Hospitalization	A0701	Capturing Problem		
									A070101	Failure to Capture		
									A070102	High Capture Threshold		
									A070104	Unstable Capture Threshold		
							F12	Serious Injury/ Illness/ Impairment	A0701	Capturing Problem		
									A070101	Failure to Capture		
									A070102	High Capture Threshold		
									A070104	Unstable Capture Threshold		
							F1905	Device Revision or Replacement	A0701	Capturing Problem		
									A070101	Failure to Capture		
									A070102	High Capture Threshold		
									A070104	Unstable Capture Threshold		
							F24	Insufficient Health Impact Information	A0701	Capturing Problem		
									A070101	Failure to Capture		
									A070102	High Capture Threshold		
									A070104	Unstable Capture Threshold		

					CT257	Implantable cardiac pacemakers	E2401	Insufficient Information	A0701	Capturing Problem		
									A070101	Failure to Capture		
									A070102	High Capture Threshold		
									A070104	Unstable Capture Threshold		
							F08	Hospitalization or Prolonged Hospitalization	A0701	Capturing Problem		
									A070101	Failure to Capture		
									A070102	High Capture Threshold		
									A070104	Unstable Capture Threshold		
							F12	Serious Injury/ Illness/ Impairment	A0701	Capturing Problem		
									A070101	Failure to Capture		
									A070102	High Capture Threshold		
									A070104	Unstable Capture Threshold		
							F1905	Device Revision or Replacement	A0701	Capturing Problem		
									A070101	Failure to Capture		
									A070102	High Capture Threshold		
									A070104	Unstable Capture Threshold		
							F24	Insufficient Health Impact Information	A0701	Capturing Problem		
									A070101	Failure to Capture		
									A070102	High Capture Threshold		
									A070104	Unstable Capture Threshold		
					CT2818	Catheterization/Cannulation/Fluid path devices	E2401	Insufficient Information	A0701	Capturing Problem		
									A070101	Failure to Capture		
									A070102	High Capture Threshold		
									A070104	Unstable Capture Threshold		
							F08	Hospitalization or Prolonged Hospitalization	A0701	Capturing Problem		
									A070101	Failure to Capture		
									A070102	High Capture Threshold		
									A070104	Unstable Capture Threshold		
							F12	Serious Injury/ Illness/ Impairment	A0701	Capturing Problem		
									A070101	Failure to Capture		
									A070102	High Capture Threshold		
									A070104	Unstable Capture Threshold		
							F1905	Device Revision or Replacement	A0701	Capturing Problem		
									A070101	Failure to Capture		
									A070102	High Capture Threshold		
									A070104	Unstable Capture Threshold		
							F24	Insufficient Health Impact Information	A0701	Capturing Problem		
									A070101	Failure to Capture		
									A070102	High Capture Threshold		
									A070104	Unstable Capture Threshold		
					CT2866	Implantable cardiac electrical stimulators	E2401	Insufficient Information	A0701	Capturing Problem		
									A070101	Failure to Capture		
									A070102	High Capture Threshold		
									A070104	Unstable Capture Threshold		
							F08	Hospitalization or Prolonged Hospitalization	A0701	Capturing Problem		
									A070101	Failure to Capture		
									A070102	High Capture Threshold		
									A070104	Unstable Capture Threshold		
							F12	Serious Injury/ Illness/ Impairment	A0701	Capturing Problem		
									A070101	Failure to Capture		
									A070102	High Capture Threshold		
									A070104	Unstable Capture Threshold		
							F1905	Device Revision or Replacement	A0701	Capturing Problem		
									A070101	Failure to Capture		
									A070102	High Capture Threshold		
									A070104	Unstable Capture Threshold		
							F24	Insufficient Health Impact Information	A0701	Capturing Problem		
									A070101	Failure to Capture		
									A070102	High Capture Threshold		
									A070104	Unstable Capture Threshold		
15625	2024	All other reportable incidents	35224	External pacemaker, epicardial pacing	CT1685	Cardiac pacemakers	E060101	Asystole	A070101	Failure to Capture		
									A0907	No Device Output		
							F1203	Life Threatening Illness or Injury	A070101	Failure to Capture		
									A0907	No Device Output		
							F23	Unexpected Medical Intervention	A070101	Failure to Capture		
									A0907	No Device Output		
15626	2024	All other reportable incidents	35224	External pacemaker, epicardial pacing	CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem		
									A0902	Display or Visual Feedback Problem		
							F23	Unexpected Medical Intervention	A0705	Battery Problem		
									A0902	Display or Visual Feedback Problem		
15627	2024		47265	Dual-chamber implantable pacemaker, rate-responsive	CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem		
							F1901	Additional Surgery	A24	No Apparent Device Problem		
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem		
							F1901	Additional Surgery	A24	No Apparent Device Problem		
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem		
							F1901	Additional Surgery	A24	No Apparent Device Problem		
15628	2024	All other reportable incidents	37265	Dual-chamber implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E210401	Shock from Patient Lead(s)	A071302	Inappropriate/Inadequate Shock/Stimulation		
									A072001	Electromagnetic Interference		
									A090801	Signal Artifact/Noise		
							F1203	Life Threatening Illness or Injury	A071302	Inappropriate/Inadequate Shock/Stimulation		
									A072001	Electromagnetic Interference		
									A090801	Signal Artifact/Noise		
15629	2024	All other reportable incidents	47267	Single-chamber implantable pacemaker, rate-responsive	CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A0908	Incorrect, Inadequate or Imprecise Result or Readings	11/06/2024	11/06/2024
							F08	Hospitalization or Prolonged Hospitalization	A0908	Incorrect, Inadequate or Imprecise Result or Readings	11/06/2024	11/06/2024
							F1905	Device Revision or Replacement	A0908	Incorrect, Inadequate or Imprecise Result or Readings	11/06/2024	11/06/2024

					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A0908	Incorrect, Inadequate or Imprecise Result or Readings	11/06/2024	11/06/2024
							F08	Hospitalization or Prolonged Hospitalization	A0908	Incorrect, Inadequate or Imprecise Result or Readings	11/06/2024	11/06/2024
							F1905	Device Revision or Replacement	A0908	Incorrect, Inadequate or Imprecise Result or Readings	11/06/2024	11/06/2024
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A0908	Incorrect, Inadequate or Imprecise Result or Readings	11/06/2024	11/06/2024
							F08	Hospitalization or Prolonged Hospitalization	A0908	Incorrect, Inadequate or Imprecise Result or Readings	11/06/2024	11/06/2024
							F1905	Device Revision or Replacement	A0908	Incorrect, Inadequate or Imprecise Result or Readings	11/06/2024	11/06/2024
15630	2024	All other reportable incidents	35852	Single-chamber implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	05/08/2024	05/08/2024
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	05/08/2024	05/08/2024
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	05/08/2024	05/08/2024
15631	2024	All other reportable incidents	47265	Dual-chamber implantable pacemaker, rate-responsive	CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A0908	Incorrect, Inadequate or Imprecise Result or Readings	06/10/2023	06/10/2023
							F08	Hospitalization or Prolonged Hospitalization	A0908	Incorrect, Inadequate or Imprecise Result or Readings	06/10/2023	06/10/2023
							F1905	Device Revision or Replacement	A0908	Incorrect, Inadequate or Imprecise Result or Readings	06/10/2023	06/10/2023
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A0908	Incorrect, Inadequate or Imprecise Result or Readings	06/10/2023	06/10/2023
							F08	Hospitalization or Prolonged Hospitalization	A0908	Incorrect, Inadequate or Imprecise Result or Readings	06/10/2023	06/10/2023
							F1905	Device Revision or Replacement	A0908	Incorrect, Inadequate or Imprecise Result or Readings	06/10/2023	06/10/2023
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A0908	Incorrect, Inadequate or Imprecise Result or Readings	06/10/2023	06/10/2023
							F08	Hospitalization or Prolonged Hospitalization	A0908	Incorrect, Inadequate or Imprecise Result or Readings	06/10/2023	06/10/2023
							F1905	Device Revision or Replacement	A0908	Incorrect, Inadequate or Imprecise Result or Readings	06/10/2023	06/10/2023
15632	2024	All other reportable incidents	35852	Single-chamber implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E210401	Shock from Patient Lead(s)	A071302	Inappropriate/Inadequate Shock/Stimulation	01/08/2024	01/08/2024
							F08	Hospitalization or Prolonged Hospitalization	A071302	Inappropriate/Inadequate Shock/Stimulation	01/08/2024	01/08/2024
							F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A071302	Inappropriate/Inadequate Shock/Stimulation	01/08/2024	01/08/2024
							F1905	Device Revision or Replacement	A071302	Inappropriate/Inadequate Shock/Stimulation	01/08/2024	01/08/2024
15633	2024	All other reportable incidents	17882	Physiologic-monitoring defibrillation system	CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A0719	Unexpected Shutdown		
							F26	No Health Consequences or Impact	A0719	Unexpected Shutdown		
15634	2024	All other reportable incidents	47265	Dual-chamber implantable pacemaker, rate-responsive	CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A071102	Failure to Interrogate		
							F26	No Health Consequences or Impact	A071102	Failure to Interrogate		
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A071102	Failure to Interrogate		
							F26	No Health Consequences or Impact	A071102	Failure to Interrogate		
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A071102	Failure to Interrogate		
							F26	No Health Consequences or Impact	A071102	Failure to Interrogate		
15635	2024		47270	Cardiac resynchronization therapy implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	F26	No Health Consequences or Impact	A0705	Battery Problem		
15636	2024	All other reportable incidents	35852	Single-chamber implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E210401	Shock from Patient Lead(s)	A0709	Device Sensing Problem		
									A0712	Pacing Problem		
									A071302	Inappropriate/Inadequate Shock/Stimulation		
							F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A0709	Device Sensing Problem		
									A0712	Pacing Problem		
									A071302	Inappropriate/Inadequate Shock/Stimulation		
							F1203	Life Threatening Illness or Injury	A0709	Device Sensing Problem		
									A0712	Pacing Problem		
									A071302	Inappropriate/Inadequate Shock/Stimulation		
							F2307	Reprogramming of Device	A0709	Device Sensing Problem		
									A0712	Pacing Problem		
									A071302	Inappropriate/Inadequate Shock/Stimulation		
15637	2024	All other reportable incidents	35852	Single-chamber implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A0508	Noise, Audible	23/08/2024	23/08/2024
									A070504	Premature Discharge of Battery	23/08/2024	23/08/2024
							F08	Hospitalization or Prolonged Hospitalization	A0508	Noise, Audible	23/08/2024	23/08/2024
									A070504	Premature Discharge of Battery	23/08/2024	23/08/2024
							F1905	Device Revision or Replacement	A0508	Noise, Audible	23/08/2024	23/08/2024
									A070504	Premature Discharge of Battery	23/08/2024	23/08/2024
15638	2024	All other reportable incidents	47263	Cardiac resynchronization therapy implantable pacemaker	CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A0908	Incorrect, Inadequate or Imprecise Result or Readings	06/08/2024	06/08/2024
							F08	Hospitalization or Prolonged Hospitalization	A0908	Incorrect, Inadequate or Imprecise Result or Readings	06/08/2024	06/08/2024
							F1905	Device Revision or Replacement	A0908	Incorrect, Inadequate or Imprecise Result or Readings	06/08/2024	06/08/2024
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A0908	Incorrect, Inadequate or Imprecise Result or Readings	06/08/2024	06/08/2024
							F08	Hospitalization or Prolonged Hospitalization	A0908	Incorrect, Inadequate or Imprecise Result or Readings	06/08/2024	06/08/2024
							F1905	Device Revision or Replacement	A0908	Incorrect, Inadequate or Imprecise Result or Readings	06/08/2024	06/08/2024
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A0908	Incorrect, Inadequate or Imprecise Result or Readings	06/08/2024	06/08/2024
							F08	Hospitalization or Prolonged Hospitalization	A0908	Incorrect, Inadequate or Imprecise Result or Readings	06/08/2024	06/08/2024
							F1905	Device Revision or Replacement	A0908	Incorrect, Inadequate or Imprecise Result or Readings	06/08/2024	06/08/2024
15639	2024	All other reportable incidents	35852	Single-chamber implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A160601	Premature Elective Replacement Indicator		
							F1905	Device Revision or Replacement	A160601	Premature Elective Replacement Indicator		
15640	2024	All other reportable incidents	47265	Dual-chamber implantable pacemaker, rate-responsive	CT1685	Cardiac pacemakers	E2402	Appropriate Clinical Signs, Symptoms and Conditions Term/Code not Available	A24	No Apparent Device Problem	02/08/2024	02/08/2024
							F08	Hospitalization or Prolonged Hospitalization	A24	No Apparent Device Problem	02/08/2024	02/08/2024
							F12	Serious Injury/ Illness/ Impairment	A24	No Apparent Device Problem	02/08/2024	02/08/2024
							F1905	Device Revision or Replacement	A24	No Apparent Device Problem	02/08/2024	02/08/2024
							F28	Appropriate Health Impact Term/Code not Available	A24	No Apparent Device Problem	02/08/2024	02/08/2024
					CT257	Implantable cardiac pacemakers	E2402	Appropriate Clinical Signs, Symptoms and Conditions Term/Code not Available	A24	No Apparent Device Problem	02/08/2024	02/08/2024
							F08	Hospitalization or Prolonged Hospitalization	A24	No Apparent Device Problem	02/08/2024	02/08/2024

							F12	Serious Injury/ Illness/ Impairment	A24	No Apparent Device Problem	02/08/2024	02/08/2024
							F1905	Device Revision or Replacement	A24	No Apparent Device Problem	02/08/2024	02/08/2024
							F28	Appropriate Health Impact Term/Code not Available	A24	No Apparent Device Problem	02/08/2024	02/08/2024
					CT2866	Implantable cardiac electrical stimulators	E2402	Appropriate Clinical Signs, Symptoms and Conditions Term/Code not Available	A24	No Apparent Device Problem	02/08/2024	02/08/2024
							F08	Hospitalization or Prolonged Hospitalization	A24	No Apparent Device Problem	02/08/2024	02/08/2024
							F12	Serious Injury/ Illness/ Impairment	A24	No Apparent Device Problem	02/08/2024	02/08/2024
							F1905	Device Revision or Replacement	A24	No Apparent Device Problem	02/08/2024	02/08/2024
							F28	Appropriate Health Impact Term/Code not Available	A24	No Apparent Device Problem	02/08/2024	02/08/2024
15641	2024	All other reportable incidents	47265	Dual-chamber implantable pacemaker, rate-responsive	CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A071102	Failure to Interrogate		
							F26	No Health Consequences or Impact	A071102	Failure to Interrogate		
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A071102	Failure to Interrogate		
							F26	No Health Consequences or Impact	A071102	Failure to Interrogate		
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A071102	Failure to Interrogate		
							F26	No Health Consequences or Impact	A071102	Failure to Interrogate		
15642	2024	All other reportable incidents	47265	Dual-chamber implantable pacemaker, rate-responsive	CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A0203	Defective Device	08/12/2023	08/12/2023
									A0908	Incorrect, Inadequate or Imprecise Result or Readings	08/12/2023	08/12/2023
							F08	Hospitalization or Prolonged Hospitalization	A0203	Defective Device	08/12/2023	08/12/2023
									A0908	Incorrect, Inadequate or Imprecise Result or Readings	08/12/2023	08/12/2023
							F1905	Device Revision or Replacement	A0203	Defective Device	08/12/2023	08/12/2023
									A0908	Incorrect, Inadequate or Imprecise Result or Readings	08/12/2023	08/12/2023
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A0203	Defective Device	08/12/2023	08/12/2023
									A0908	Incorrect, Inadequate or Imprecise Result or Readings	08/12/2023	08/12/2023
							F08	Hospitalization or Prolonged Hospitalization	A0203	Defective Device	08/12/2023	08/12/2023
									A0908	Incorrect, Inadequate or Imprecise Result or Readings	08/12/2023	08/12/2023
							F1905	Device Revision or Replacement	A0203	Defective Device	08/12/2023	08/12/2023
									A0908	Incorrect, Inadequate or Imprecise Result or Readings	08/12/2023	08/12/2023
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A0203	Defective Device	08/12/2023	08/12/2023
									A0908	Incorrect, Inadequate or Imprecise Result or Readings	08/12/2023	08/12/2023
							F08	Hospitalization or Prolonged Hospitalization	A0203	Defective Device	08/12/2023	08/12/2023
									A0908	Incorrect, Inadequate or Imprecise Result or Readings	08/12/2023	08/12/2023
							F1905	Device Revision or Replacement	A0203	Defective Device	08/12/2023	08/12/2023
									A0908	Incorrect, Inadequate or Imprecise Result or Readings	08/12/2023	08/12/2023
15643	2024	All other reportable incidents	47270	Cardiac resynchronization therapy implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A071102	Failure to Interrogate		
							F26	No Health Consequences or Impact	A071102	Failure to Interrogate		
15644	2024	All other reportable incidents	37265	Dual-chamber implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A071102	Failure to Interrogate		
							F26	No Health Consequences or Impact	A071102	Failure to Interrogate		
15645	2024	All other reportable incidents	47267	Single-chamber implantable pacemaker, rate-responsive	CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	17/07/2024	17/07/2024
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery	17/07/2024	17/07/2024
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	17/07/2024	17/07/2024
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery	17/07/2024	17/07/2024
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	17/07/2024	17/07/2024
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery	17/07/2024	17/07/2024
15646	2024	All other reportable incidents	47263	Cardiac resynchronization therapy implantable pacemaker	CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A071102	Failure to Interrogate		
							F26	No Health Consequences or Impact	A071102	Failure to Interrogate		
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A071102	Failure to Interrogate		
							F26	No Health Consequences or Impact	A071102	Failure to Interrogate		
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A071102	Failure to Interrogate		
							F26	No Health Consequences or Impact	A071102	Failure to Interrogate		
15647	2024		47265	Dual-chamber implantable pacemaker, rate-responsive	CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	13/07/2024	13/07/2024
							F1901	Additional Surgery	A24	No Apparent Device Problem	13/07/2024	13/07/2024
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	13/07/2024	13/07/2024
							F1901	Additional Surgery	A24	No Apparent Device Problem	13/07/2024	13/07/2024
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	13/07/2024	13/07/2024
							F1901	Additional Surgery	A24	No Apparent Device Problem	13/07/2024	13/07/2024
15648	2024	All other reportable incidents	47265	Dual-chamber implantable pacemaker, rate-responsive	CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	29/07/2024	29/07/2024
							F1901	Additional Surgery	A24	No Apparent Device Problem	29/07/2024	29/07/2024
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	29/07/2024	29/07/2024
							F1901	Additional Surgery	A24	No Apparent Device Problem	29/07/2024	29/07/2024
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	29/07/2024	29/07/2024
							F1901	Additional Surgery	A24	No Apparent Device Problem	29/07/2024	29/07/2024
15649	2024	Unanticipated serious deterioration in state of health	47265	Dual-chamber implantable pacemaker, rate-responsive	CT1685	Cardiac pacemakers	E011903	Syncope/Fainting	A071102	Failure to Interrogate		
									A0712	Pacing Problem		
							E0618	Non Specific EKG/ECG Changes	A071102	Failure to Interrogate		
									A0712	Pacing Problem		
							E2312	Fatigue	A071102	Failure to Interrogate		
									A0712	Pacing Problem		
							F1203	Life Threatening Illness or Injury	A071102	Failure to Interrogate		
									A0712	Pacing Problem		
							F2307	Reprogramming of Device	A071102	Failure to Interrogate		
									A0712	Pacing Problem		
					CT257	Implantable cardiac pacemakers	E011903	Syncope/Fainting	A071102	Failure to Interrogate		
									A0712	Pacing Problem		
							E0618	Non Specific EKG/ECG Changes	A071102	Failure to Interrogate		
									A0712	Pacing Problem		
							E2312	Fatigue	A071102	Failure to Interrogate		
									A0712	Pacing Problem		
							F1203	Life Threatening Illness or Injury	A071102	Failure to Interrogate		
									A0712	Pacing Problem		
							F2307	Reprogramming of Device	A071102	Failure to Interrogate		
									A0712	Pacing Problem		

					CT2866	Implantable cardiac electrical stimulators	E011903	Syncope/Fainting	A071102	Failure to Interrogate		
									A0712	Pacing Problem		
							E0618	Non Specific EKG/ECG Changes	A071102	Failure to Interrogate		
									A0712	Pacing Problem		
							E2312	Fatigue	A071102	Failure to Interrogate		
									A0712	Pacing Problem		
							F1203	Life Threatening Illness or Injury	A071102	Failure to Interrogate		
									A0712	Pacing Problem		
							F2307	Reprogramming of Device	A071102	Failure to Interrogate		
									A0712	Pacing Problem		
15650	2024	All other reportable incidents	47265	Dual-chamber implantable pacemaker, rate-responsive	CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A0508	Noise, Audible	05/08/2024	05/08/2024
									A0712	Pacing Problem	05/08/2024	05/08/2024
							F26	No Health Consequences or Impact	A0508	Noise, Audible	05/08/2024	05/08/2024
									A0712	Pacing Problem	05/08/2024	05/08/2024
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A0508	Noise, Audible	05/08/2024	05/08/2024
									A0712	Pacing Problem	05/08/2024	05/08/2024
							F26	No Health Consequences or Impact	A0508	Noise, Audible	05/08/2024	05/08/2024
									A0712	Pacing Problem	05/08/2024	05/08/2024
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A0508	Noise, Audible	05/08/2024	05/08/2024
									A0712	Pacing Problem	05/08/2024	05/08/2024
							F26	No Health Consequences or Impact	A0508	Noise, Audible	05/08/2024	05/08/2024
									A0712	Pacing Problem	05/08/2024	05/08/2024
15651	2024	All other reportable incidents	17882	Physiologic-monitoring defibrillation system	CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070908	Failure to Sense		
							F26	No Health Consequences or Impact	A070908	Failure to Sense		
15652	2024	All other reportable incidents	47265	Dual-chamber implantable pacemaker, rate-responsive	CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A0908	Incorrect, Inadequate or Imprecise Result or Readings	14/08/2024	14/08/2024
							F08	Hospitalization or Prolonged Hospitalization	A0908	Incorrect, Inadequate or Imprecise Result or Readings	14/08/2024	14/08/2024
							F1905	Device Revision or Replacement	A0908	Incorrect, Inadequate or Imprecise Result or Readings	14/08/2024	14/08/2024
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A0908	Incorrect, Inadequate or Imprecise Result or Readings	14/08/2024	14/08/2024
							F08	Hospitalization or Prolonged Hospitalization	A0908	Incorrect, Inadequate or Imprecise Result or Readings	14/08/2024	14/08/2024
							F1905	Device Revision or Replacement	A0908	Incorrect, Inadequate or Imprecise Result or Readings	14/08/2024	14/08/2024
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A0908	Incorrect, Inadequate or Imprecise Result or Readings	14/08/2024	14/08/2024
							F08	Hospitalization or Prolonged Hospitalization	A0908	Incorrect, Inadequate or Imprecise Result or Readings	14/08/2024	14/08/2024
							F1905	Device Revision or Replacement	A0908	Incorrect, Inadequate or Imprecise Result or Readings	14/08/2024	14/08/2024
15653	2024	Death	47270	Cardiac resynchronization therapy implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E0611	Heart Failure/Congestive Heart Failure	A070504	Premature Discharge of Battery	05/08/2024	05/08/2024
							E1014	Gastrointestinal Hemorrhage	A070504	Premature Discharge of Battery	05/08/2024	05/08/2024
							F02	Death	A070504	Premature Discharge of Battery	05/08/2024	05/08/2024
15654	2024	All other reportable incidents	35852	Single-chamber implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A0709	Device Sensing Problem		
									A090801	Signal Artifact/Noise		
							F12	Serious Injury/ Illness/ Impairment	A0709	Device Sensing Problem		
									A090801	Signal Artifact/Noise		
							F2307	Reprogramming of Device	A0709	Device Sensing Problem		
									A090801	Signal Artifact/Noise		
15655	2024	All other reportable incidents	36241	Pacemaker/defibrillator lead	CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070102	High Capture Threshold		
									A072201	High Impedance		
									A090801	Signal Artifact/Noise		
							F26	No Health Consequences or Impact	A070102	High Capture Threshold		
									A072201	High Impedance		
									A090801	Signal Artifact/Noise		
15656	2024	All other reportable incidents	47265	Dual-chamber implantable pacemaker, rate-responsive	CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A070102	High Capture Threshold		
									A072201	High Impedance		
									A090801	Signal Artifact/Noise		
							F26	No Health Consequences or Impact	A070102	High Capture Threshold		
									A072201	High Impedance		
									A090801	Signal Artifact/Noise		
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A070102	High Capture Threshold		
									A072201	High Impedance		
									A090801	Signal Artifact/Noise		
							F26	No Health Consequences or Impact	A070102	High Capture Threshold		
									A072201	High Impedance		
									A090801	Signal Artifact/Noise		
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A070102	High Capture Threshold		
									A072201	High Impedance		
									A090801	Signal Artifact/Noise		
							F26	No Health Consequences or Impact	A070102	High Capture Threshold		
									A072201	High Impedance		
									A090801	Signal Artifact/Noise		
15657	2024	All other reportable incidents	47205	Cardiac pulse generator programmer	CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A0902	Display or Visual Feedback Problem		
							F27	Problem Identified During Non-Clinical Procedure	A0902	Display or Visual Feedback Problem		
15658	2024	All other reportable incidents	17882	Physiologic-monitoring defibrillation system	CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070802	Intermittent Loss of Power		
							F27	Problem Identified During Non-Clinical Procedure	A070802	Intermittent Loss of Power		
15659	2024	All other reportable incidents	47265	Dual-chamber implantable pacemaker, rate-responsive	CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A0709	Device Sensing Problem	24/05/2024	24/05/2024
									A0712	Pacing Problem	24/05/2024	24/05/2024
							F08	Hospitalization or Prolonged Hospitalization	A0709	Device Sensing Problem	24/05/2024	24/05/2024
									A0712	Pacing Problem	24/05/2024	24/05/2024
							F1203	Life Threatening Illness or Injury	A0709	Device Sensing Problem	24/05/2024	24/05/2024
									A0712	Pacing Problem	24/05/2024	24/05/2024
							F1905	Device Revision or Replacement	A0709	Device Sensing Problem	24/05/2024	24/05/2024
									A0712	Pacing Problem	24/05/2024	24/05/2024
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A0709	Device Sensing Problem	24/05/2024	24/05/2024
									A0712	Pacing Problem	24/05/2024	24/05/2024
							F08	Hospitalization or Prolonged Hospitalization	A0709	Device Sensing Problem	24/05/2024	24/05/2024

									A0712	Pacing Problem	24/05/2024	24/05/2024
							F1203	Life Threatening Illness or Injury	A0709	Device Sensing Problem	24/05/2024	24/05/2024
									A0712	Pacing Problem	24/05/2024	24/05/2024
							F1905	Device Revision or Replacement	A0709	Device Sensing Problem	24/05/2024	24/05/2024
									A0712	Pacing Problem	24/05/2024	24/05/2024
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A0709	Device Sensing Problem	24/05/2024	24/05/2024
									A0712	Pacing Problem	24/05/2024	24/05/2024
							F08	Hospitalization or Prolonged Hospitalization	A0709	Device Sensing Problem	24/05/2024	24/05/2024
									A0712	Pacing Problem	24/05/2024	24/05/2024
							F1203	Life Threatening Illness or Injury	A0709	Device Sensing Problem	24/05/2024	24/05/2024
									A0712	Pacing Problem	24/05/2024	24/05/2024
							F1905	Device Revision or Replacement	A0709	Device Sensing Problem	24/05/2024	24/05/2024
									A0712	Pacing Problem	24/05/2024	24/05/2024
15660	2024	All other reportable incidents	37265	Dual-chamber implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A05	Mechanical Problem	19/07/2024	19/07/2024
									A090801	Signal Artifact/Noise	19/07/2024	19/07/2024
							F26	No Health Consequences or Impact	A05	Mechanical Problem	19/07/2024	19/07/2024
									A090801	Signal Artifact/Noise	19/07/2024	19/07/2024
15661	2024	All other reportable incidents	47263	Cardiac resynchronization therapy implantable pacemaker	CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A0908	Incorrect, Inadequate or Imprecise Result or Readings	26/07/2024	26/07/2024
							F08	Hospitalization or Prolonged Hospitalization	A0908	Incorrect, Inadequate or Imprecise Result or Readings	26/07/2024	26/07/2024
							F1905	Device Revision or Replacement	A0908	Incorrect, Inadequate or Imprecise Result or Readings	26/07/2024	26/07/2024
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A0908	Incorrect, Inadequate or Imprecise Result or Readings	26/07/2024	26/07/2024
							F08	Hospitalization or Prolonged Hospitalization	A0908	Incorrect, Inadequate or Imprecise Result or Readings	26/07/2024	26/07/2024
							F1905	Device Revision or Replacement	A0908	Incorrect, Inadequate or Imprecise Result or Readings	26/07/2024	26/07/2024
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A0908	Incorrect, Inadequate or Imprecise Result or Readings	26/07/2024	26/07/2024
							F08	Hospitalization or Prolonged Hospitalization	A0908	Incorrect, Inadequate or Imprecise Result or Readings	26/07/2024	26/07/2024
							F1905	Device Revision or Replacement	A0908	Incorrect, Inadequate or Imprecise Result or Readings	26/07/2024	26/07/2024
15662	2024	All other reportable incidents	17882	Physiologic-monitoring defibrillation system	CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A07	Electrical /Electronic Property Problem		
							F27	Problem Identified During Non-Clinical Procedure	A07	Electrical /Electronic Property Problem		
15663	2024	All other reportable incidents	17882	Physiologic-monitoring defibrillation system	CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070802	Intermittent Loss of Power		
							F27	Problem Identified During Non-Clinical Procedure	A070802	Intermittent Loss of Power		
15664	2024	All other reportable incidents	17882	Physiologic-monitoring defibrillation system	CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A0719	Unexpected Shutdown		
							F27	Problem Identified During Non-Clinical Procedure	A0719	Unexpected Shutdown		
15665	2024	All other reportable incidents	47267	Single-chamber implantable pacemaker, rate-responsive	CT1685	Cardiac pacemakers	E0112	Dizziness	A070101	Failure to Capture		
									A070909	Oversensing		
									A071206	Pacing Asynchronously		
							F11	Non-Serious Injury/ Illness/ Impairment	A070101	Failure to Capture		
									A070909	Oversensing		
									A071206	Pacing Asynchronously		
					CT257	Implantable cardiac pacemakers	E0112	Dizziness	A070101	Failure to Capture		
									A070909	Oversensing		
									A071206	Pacing Asynchronously		
							F11	Non-Serious Injury/ Illness/ Impairment	A070101	Failure to Capture		
									A070909	Oversensing		
									A071206	Pacing Asynchronously		
					CT2866	Implantable cardiac electrical stimulators	E0112	Dizziness	A070101	Failure to Capture		
									A070909	Oversensing		
									A071206	Pacing Asynchronously		
							F11	Non-Serious Injury/ Illness/ Impairment	A070101	Failure to Capture		
									A070909	Oversensing		
									A071206	Pacing Asynchronously		
15666	2024	All other reportable incidents	17882	Physiologic-monitoring defibrillation system	CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A0719	Unexpected Shutdown		
							F27	Problem Identified During Non-Clinical Procedure	A0719	Unexpected Shutdown		
15667	2024		47265	Dual-chamber implantable pacemaker, rate-responsive	CT1685	Cardiac pacemakers	E020204	Malaise	A0203	Defective Device		
									A071205	Pacemaker Found in Back-Up Mode		
							E2007	Fall	A0203	Defective Device		
									A071205	Pacemaker Found in Back-Up Mode		
							F12	Serious Injury/ Illness/ Impairment	A0203	Defective Device		
									A071205	Pacemaker Found in Back-Up Mode		
					CT257	Implantable cardiac pacemakers	E020204	Malaise	A0203	Defective Device		
									A071205	Pacemaker Found in Back-Up Mode		
							E2007	Fall	A0203	Defective Device		
									A071205	Pacemaker Found in Back-Up Mode		
							F12	Serious Injury/ Illness/ Impairment	A0203	Defective Device		
									A071205	Pacemaker Found in Back-Up Mode		
					CT2866	Implantable cardiac electrical stimulators	E020204	Malaise	A0203	Defective Device		
									A071205	Pacemaker Found in Back-Up Mode		
							E2007	Fall	A0203	Defective Device		
									A071205	Pacemaker Found in Back-Up Mode		
							F12	Serious Injury/ Illness/ Impairment	A0203	Defective Device		
									A071205	Pacemaker Found in Back-Up Mode		
15668	2024	Death	17882	Physiologic-monitoring defibrillation system	CT1684	Cardiac pacemakers and associated devices	E0601	Arrhythmia	A0707	Failure to Discharge		
							F02	Death	A0707	Failure to Discharge		
15669	2024	All other reportable incidents	47265	Dual-chamber implantable pacemaker, rate-responsive	CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	22/10/2024	22/10/2024
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	22/10/2024	22/10/2024
							F1903	Device Explantation	A070504	Premature Discharge of Battery	22/10/2024	22/10/2024
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	22/10/2024	22/10/2024
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	22/10/2024	22/10/2024
							F1903	Device Explantation	A070504	Premature Discharge of Battery	22/10/2024	22/10/2024
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	22/10/2024	22/10/2024
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	22/10/2024	22/10/2024
							F1903	Device Explantation	A070504	Premature Discharge of Battery	22/10/2024	22/10/2024

15670	2024	All other reportable incidents	47265	Dual-chamber implantable pacemaker, rate-responsive	CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	25/01/2024	25/01/2024
									A0908	Incorrect, Inadequate or Imprecise Result or Readings	25/01/2024	25/01/2024
									A1304	Telemetry Discrepancy	25/01/2024	25/01/2024
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	25/01/2024	25/01/2024
									A0908	Incorrect, Inadequate or Imprecise Result or Readings	25/01/2024	25/01/2024
									A1304	Telemetry Discrepancy	25/01/2024	25/01/2024
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	25/01/2024	25/01/2024
									A0908	Incorrect, Inadequate or Imprecise Result or Readings	25/01/2024	25/01/2024
									A1304	Telemetry Discrepancy	25/01/2024	25/01/2024
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	25/01/2024	25/01/2024
									A0908	Incorrect, Inadequate or Imprecise Result or Readings	25/01/2024	25/01/2024
									A1304	Telemetry Discrepancy	25/01/2024	25/01/2024
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	25/01/2024	25/01/2024
									A0908	Incorrect, Inadequate or Imprecise Result or Readings	25/01/2024	25/01/2024
									A1304	Telemetry Discrepancy	25/01/2024	25/01/2024
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	25/01/2024	25/01/2024
									A0908	Incorrect, Inadequate or Imprecise Result or Readings	25/01/2024	25/01/2024
									A1304	Telemetry Discrepancy	25/01/2024	25/01/2024
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	25/01/2024	25/01/2024
									A0908	Incorrect, Inadequate or Imprecise Result or Readings	25/01/2024	25/01/2024
									A1304	Telemetry Discrepancy	25/01/2024	25/01/2024
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	25/01/2024	25/01/2024
									A0908	Incorrect, Inadequate or Imprecise Result or Readings	25/01/2024	25/01/2024
									A1304	Telemetry Discrepancy	25/01/2024	25/01/2024
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	25/01/2024	25/01/2024
									A0908	Incorrect, Inadequate or Imprecise Result or Readings	25/01/2024	25/01/2024
									A1304	Telemetry Discrepancy	25/01/2024	25/01/2024
15671	2024	All other reportable incidents	35852	Single-chamber implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E210401	Shock from Patient Lead(s)	A0709	Device Sensing Problem		
									A071302	Inappropriate/Inadequate Shock/Stimulation		
							F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A0709	Device Sensing Problem		
									A071302	Inappropriate/Inadequate Shock/Stimulation		
							F1203	Life Threatening Illness or Injury	A0709	Device Sensing Problem		
									A071302	Inappropriate/Inadequate Shock/Stimulation		
15672	2024	All other reportable incidents	47267	Single-chamber implantable pacemaker, rate-responsive	CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A070102	High Capture Threshold	07/08/2024	07/08/2024
									A070504	Premature Discharge of Battery	07/08/2024	07/08/2024
									A160602	Premature End-of-Life Indicator	07/08/2024	07/08/2024
							F08	Hospitalization or Prolonged Hospitalization	A070102	High Capture Threshold	07/08/2024	07/08/2024
									A070504	Premature Discharge of Battery	07/08/2024	07/08/2024
									A160602	Premature End-of-Life Indicator	07/08/2024	07/08/2024
							F1905	Device Revision or Replacement	A070102	High Capture Threshold	07/08/2024	07/08/2024
									A070504	Premature Discharge of Battery	07/08/2024	07/08/2024
									A160602	Premature End-of-Life Indicator	07/08/2024	07/08/2024
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A070102	High Capture Threshold	07/08/2024	07/08/2024
									A070504	Premature Discharge of Battery	07/08/2024	07/08/2024
									A160602	Premature End-of-Life Indicator	07/08/2024	07/08/2024
							F08	Hospitalization or Prolonged Hospitalization	A070102	High Capture Threshold	07/08/2024	07/08/2024
									A070504	Premature Discharge of Battery	07/08/2024	07/08/2024
									A160602	Premature End-of-Life Indicator	07/08/2024	07/08/2024
							F1905	Device Revision or Replacement	A070102	High Capture Threshold	07/08/2024	07/08/2024
									A070504	Premature Discharge of Battery	07/08/2024	07/08/2024
									A160602	Premature End-of-Life Indicator	07/08/2024	07/08/2024
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A070102	High Capture Threshold	07/08/2024	07/08/2024
									A070504	Premature Discharge of Battery	07/08/2024	07/08/2024
									A160602	Premature End-of-Life Indicator	07/08/2024	07/08/2024
							F08	Hospitalization or Prolonged Hospitalization	A070102	High Capture Threshold	07/08/2024	07/08/2024
									A070504	Premature Discharge of Battery	07/08/2024	07/08/2024
									A160602	Premature End-of-Life Indicator	07/08/2024	07/08/2024
							F1905	Device Revision or Replacement	A070102	High Capture Threshold	07/08/2024	07/08/2024
									A070504	Premature Discharge of Battery	07/08/2024	07/08/2024
									A160602	Premature End-of-Life Indicator	07/08/2024	07/08/2024
15673	2024	All other reportable incidents	47265	Dual-chamber implantable pacemaker, rate-responsive	CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
15674	2024	All other reportable incidents	47270	Cardiac resynchronization therapy implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	13/08/2024	13/08/2024
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery	13/08/2024	13/08/2024
15675	2024	All other reportable incidents	36106	Implantable pacemaker analyser, intraoperative	CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A0902	Display or Visual Feedback Problem		
									A11	Computer Software Problem		
							F26	No Health Consequences or Impact	A0902	Display or Visual Feedback Problem		
									A11	Computer Software Problem		
15676	2024	All other reportable incidents	17882	Physiologic-monitoring defibrillation system	CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A25	No Apparent Adverse Event		
							F27	Problem Identified During Non-Clinical Procedure	A25	No Apparent Adverse Event		
15677	2024	All other reportable incidents	47265	Dual-chamber implantable pacemaker, rate-responsive	CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem	31/07/2024	31/07/2024
									A070504	Premature Discharge of Battery	31/07/2024	31/07/2024
							F08	Hospitalization or Prolonged Hospitalization	A0705	Battery Problem	31/07/2024	31/07/2024
									A070504	Premature Discharge of Battery	31/07/2024	31/07/2024
							F1905	Device Revision or Replacement	A0705	Battery Problem	31/07/2024	31/07/2024
									A070504	Premature Discharge of Battery	31/07/2024	31/07/2024
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem	31/07/2024	31/07/2024

									A070504	Premature Discharge of Battery	31/07/2024	31/07/2024
							F08	Hospitalization or Prolonged Hospitalization	A0705	Battery Problem	31/07/2024	31/07/2024
									A070504	Premature Discharge of Battery	31/07/2024	31/07/2024
							F1905	Device Revision or Replacement	A0705	Battery Problem	31/07/2024	31/07/2024
									A070504	Premature Discharge of Battery	31/07/2024	31/07/2024
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem	31/07/2024	31/07/2024
									A070504	Premature Discharge of Battery	31/07/2024	31/07/2024
							F08	Hospitalization or Prolonged Hospitalization	A0705	Battery Problem	31/07/2024	31/07/2024
									A070504	Premature Discharge of Battery	31/07/2024	31/07/2024
							F1905	Device Revision or Replacement	A0705	Battery Problem	31/07/2024	31/07/2024
									A070504	Premature Discharge of Battery	31/07/2024	31/07/2024
15678	2024	All other reportable incidents	17882	Physiologic-monitoring defibrillation system	CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070803	Failure to Power Up		
							F27	Problem Identified During Non-Clinical Procedure	A070803	Failure to Power Up		
15679	2024	All other reportable incidents	47265	Dual-chamber implantable pacemaker, rate-responsive	CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	09/09/2024	09/09/2024
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	09/09/2024	09/09/2024
							F1903	Device Explantation	A070504	Premature Discharge of Battery	09/09/2024	09/09/2024
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	09/09/2024	09/09/2024
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	09/09/2024	09/09/2024
							F1903	Device Explantation	A070504	Premature Discharge of Battery	09/09/2024	09/09/2024
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	09/09/2024	09/09/2024
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	09/09/2024	09/09/2024
							F1903	Device Explantation	A070504	Premature Discharge of Battery	09/09/2024	09/09/2024
15680	2024	All other reportable incidents	47265	Dual-chamber implantable pacemaker, rate-responsive	CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	06/08/2024	06/08/2024
							F1901	Additional Surgery	A24	No Apparent Device Problem	06/08/2024	06/08/2024
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	06/08/2024	06/08/2024
							F1901	Additional Surgery	A24	No Apparent Device Problem	06/08/2024	06/08/2024
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	06/08/2024	06/08/2024
							F1901	Additional Surgery	A24	No Apparent Device Problem	06/08/2024	06/08/2024
15681	2024		35822	External pacemaker, transcutaneous pacing	CT1685	Cardiac pacemakers	E0602	Cardiac Arrest	A0203	Defective Device		
							F1203	Life Threatening Illness or Injury	A0203	Defective Device		
15682	2024		47265	Dual-chamber implantable pacemaker, rate-responsive	CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A0203	Defective Device		
									A07	Electrical /Electronic Property Problem		
									A0705	Battery Problem		
									A160601	Premature Elective Replacement Indicator		
							F1905	Device Revision or Replacement	A0203	Defective Device		
									A07	Electrical /Electronic Property Problem		
									A0705	Battery Problem		
									A160601	Premature Elective Replacement Indicator		
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A0203	Defective Device		
									A07	Electrical /Electronic Property Problem		
									A0705	Battery Problem		
									A160601	Premature Elective Replacement Indicator		
							F1905	Device Revision or Replacement	A0203	Defective Device		
									A07	Electrical /Electronic Property Problem		
									A0705	Battery Problem		
									A160601	Premature Elective Replacement Indicator		
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A0203	Defective Device		
									A07	Electrical /Electronic Property Problem		
									A0705	Battery Problem		
									A160601	Premature Elective Replacement Indicator		
							F1905	Device Revision or Replacement	A0203	Defective Device		
									A07	Electrical /Electronic Property Problem		
									A0705	Battery Problem		
									A160601	Premature Elective Replacement Indicator		
15683	2024	All other reportable incidents	47265	Dual-chamber implantable pacemaker, rate-responsive	CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
15684	2024	All other reportable incidents	47265	Dual-chamber implantable pacemaker, rate-responsive	CT141	Electrodes/Leads and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT145	Cardiovascular devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT233	Surgical	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT2406	Implantable	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT2407	Active implantable	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT2727	Active implant devices/ancillaries	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT2750	Implantable cardiac electrical stimulators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT2985	Active	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT3001	Implantable metal	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT3002	Implantable synthetic polymer	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		

					CT3006	Implantable inorganic material	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT314	Electrically powered	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT321	Long-term surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT449	Generators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT981	Single use	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT983	Surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
15685	2024	All other reportable incidents	47265	Dual-chamber implantable pacemaker, rate-responsive	CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
15686	2024	All other reportable incidents	47940	Triple-chamber implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A070102	High Capture Threshold		
							F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A070102	High Capture Threshold		
15687	2024	All other reportable incidents	17882	Physiologic-monitoring defibrillation system	CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070802	Intermittent Loss of Power		
							F27	Problem Identified During Non-Clinical Procedure	A070802	Intermittent Loss of Power		
15688	2024	All other reportable incidents	17882	Physiologic-monitoring defibrillation system	CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070802	Intermittent Loss of Power		
							F27	Problem Identified During Non-Clinical Procedure	A070802	Intermittent Loss of Power		
15689	2024	All other reportable incidents	47265	Dual-chamber implantable pacemaker, rate-responsive	CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A070101	Failure to Capture		
							F26	No Health Consequences or Impact	A070101	Failure to Capture		
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A070101	Failure to Capture		
							F26	No Health Consequences or Impact	A070101	Failure to Capture		
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A070101	Failure to Capture		
							F26	No Health Consequences or Impact	A070101	Failure to Capture		
15690	2024	All other reportable incidents	47205	Cardiac pulse generator programmer	CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A0902	Display or Visual Feedback Problem		
							F27	Problem Identified During Non-Clinical Procedure	A0902	Display or Visual Feedback Problem		
15691	2024	All other reportable incidents	17882	Physiologic-monitoring defibrillation system	CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A071301	Failure to Deliver Shock/Stimulation		
							F26	No Health Consequences or Impact	A071301	Failure to Deliver Shock/Stimulation		
15692	2024	All other reportable incidents	37265	Dual-chamber implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A070909	Oversensing		
									A072001	Electromagnetic Interference		
									A090801	Signal Artifact/Noise		
							F26	No Health Consequences or Impact	A070909	Oversensing		
									A072001	Electromagnetic Interference		
									A090801	Signal Artifact/Noise		
15693	2024	All other reportable incidents	47270	Cardiac resynchronization therapy implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E210401	Shock from Patient Lead(s)	A0709	Device Sensing Problem		
									A071302	Inappropriate/Inadequate Shock/Stimulation		
							F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A0709	Device Sensing Problem		
									A071302	Inappropriate/Inadequate Shock/Stimulation		
							F1203	Life Threatening Illness or Injury	A0709	Device Sensing Problem		
									A071302	Inappropriate/Inadequate Shock/Stimulation		
							F2307	Reprogramming of Device	A0709	Device Sensing Problem		
									A071302	Inappropriate/Inadequate Shock/Stimulation		
15694	2024	All other reportable incidents	47205	Cardiac pulse generator programmer	CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A0902	Display or Visual Feedback Problem		
							F27	Problem Identified During Non-Clinical Procedure	A0902	Display or Visual Feedback Problem		
15695	2024	Unanticipated serious deterioration in state of health	47265	Dual-chamber implantable pacemaker, rate-responsive	CT1685	Cardiac pacemakers	E0402	Hypersensitivity/Allergic Reaction	A24	No Apparent Device Problem	01/12/2015	31/12/2015
							F08	Hospitalization or Prolonged Hospitalization	A24	No Apparent Device Problem	01/12/2015	31/12/2015
							F1905	Device Revision or Replacement	A24	No Apparent Device Problem	01/12/2015	31/12/2015
					CT257	Implantable cardiac pacemakers	E0402	Hypersensitivity/Allergic Reaction	A24	No Apparent Device Problem	01/12/2015	31/12/2015
							F08	Hospitalization or Prolonged Hospitalization	A24	No Apparent Device Problem	01/12/2015	31/12/2015
							F1905	Device Revision or Replacement	A24	No Apparent Device Problem	01/12/2015	31/12/2015
					CT2866	Implantable cardiac electrical stimulators	E0402	Hypersensitivity/Allergic Reaction	A24	No Apparent Device Problem	01/12/2015	31/12/2015
							F08	Hospitalization or Prolonged Hospitalization	A24	No Apparent Device Problem	01/12/2015	31/12/2015
							F1905	Device Revision or Replacement	A24	No Apparent Device Problem	01/12/2015	31/12/2015
15696	2024		17882	Physiologic-monitoring defibrillation system	CT1684	Cardiac pacemakers and associated devices	E2401	Insufficient Information	A070908	Failure to Sense		
							F2301	Additional Device Required	A070908	Failure to Sense		
15697	2024	All other reportable incidents	35852	Single-chamber implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A1605	Reset Problem		
							F26	No Health Consequences or Impact	A1605	Reset Problem		
15698	2024	All other reportable incidents	35852	Single-chamber implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A1605	Reset Problem		
							F26	No Health Consequences or Impact	A1605	Reset Problem		
15699	2024	All other reportable incidents	37265	Dual-chamber implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E210401	Shock from Patient Lead(s)	A071302	Inappropriate/Inadequate Shock/Stimulation		
							F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A071302	Inappropriate/Inadequate Shock/Stimulation		
15700	2024	All other reportable incidents	16038	Implantable pulse generator mesh bag, non-bioabsorbable	CT1684	Cardiac pacemakers and associated devices	E210401	Shock from Patient Lead(s)	A071302	Inappropriate/Inadequate Shock/Stimulation		
							F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A071302	Inappropriate/Inadequate Shock/Stimulation		
							F2203	Imaging Required	A071302	Inappropriate/Inadequate Shock/Stimulation		
15701	2024	All other reportable incidents	17882	Physiologic-monitoring defibrillation system	CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A071303	Intermittent Shock/Stimulation		
							F27	Problem Identified During Non-Clinical Procedure	A071303	Intermittent Shock/Stimulation		
15702	2024	All other reportable incidents	47205	Cardiac pulse generator programmer	CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A0902	Display or Visual Feedback Problem		
							F27	Problem Identified During Non-Clinical Procedure	A0902	Display or Visual Feedback Problem		
15703	2024	All other reportable incidents	47205	Cardiac pulse generator programmer	CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A0902	Display or Visual Feedback Problem		
							F27	Problem Identified During Non-Clinical Procedure	A0902	Display or Visual Feedback Problem		
15704	2024	All other reportable incidents	47205	Cardiac pulse generator programmer	CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A0902	Display or Visual Feedback Problem		
									A11	Computer Software Problem		
									A1104	Computer Operating System Problem		
							F27	Problem Identified During Non-Clinical Procedure	A0902	Display or Visual Feedback Problem		
									A11	Computer Software Problem		
									A1104	Computer Operating System Problem		
15705	2024	All other reportable incidents	37265	Dual-chamber implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
15706	2024	All other reportable incidents	47263	Cardiac resynchronization therapy implantable pacemaker	CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A0908	Incorrect, Inadequate or Imprecise Result or Readings	06/08/2024	06/08/2024
							F08	Hospitalization or Prolonged Hospitalization	A0908	Incorrect, Inadequate or Imprecise Result or Readings	06/08/2024	06/08/2024

							F1905	Device Revision or Replacement	A0908	Incorrect, Inadequate or Imprecise Result or Readings	06/08/2024	06/08/2024
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A0908	Incorrect, Inadequate or Imprecise Result or Readings	06/08/2024	06/08/2024
							F08	Hospitalization or Prolonged Hospitalization	A0908	Incorrect, Inadequate or Imprecise Result or Readings	06/08/2024	06/08/2024
							F1905	Device Revision or Replacement	A0908	Incorrect, Inadequate or Imprecise Result or Readings	06/08/2024	06/08/2024
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A0908	Incorrect, Inadequate or Imprecise Result or Readings	06/08/2024	06/08/2024
							F08	Hospitalization or Prolonged Hospitalization	A0908	Incorrect, Inadequate or Imprecise Result or Readings	06/08/2024	06/08/2024
							F1905	Device Revision or Replacement	A0908	Incorrect, Inadequate or Imprecise Result or Readings	06/08/2024	06/08/2024
15707	2024	All other reportable incidents	47267	Single-chamber implantable pacemaker, rate-responsive	CT1685	Cardiac pacemakers	E2311	Discomfort	A0723	Unexpected Mode Switch		
									A090407	Therapy Delivered to Incorrect Body Area		
							F23	Unexpected Medical Intervention	A0723	Unexpected Mode Switch		
									A090407	Therapy Delivered to Incorrect Body Area		
					CT257	Implantable cardiac pacemakers	E2311	Discomfort	A0723	Unexpected Mode Switch		
									A090407	Therapy Delivered to Incorrect Body Area		
							F23	Unexpected Medical Intervention	A0723	Unexpected Mode Switch		
									A090407	Therapy Delivered to Incorrect Body Area		
					CT2866	Implantable cardiac electrical stimulators	E2311	Discomfort	A0723	Unexpected Mode Switch		
									A090407	Therapy Delivered to Incorrect Body Area		
							F23	Unexpected Medical Intervention	A0723	Unexpected Mode Switch		
									A090407	Therapy Delivered to Incorrect Body Area		
15708	2024		47265	Dual-chamber implantable pacemaker, rate-responsive	CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	18/03/2024	18/03/2024
							F1901	Additional Surgery	A24	No Apparent Device Problem	18/03/2024	18/03/2024
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	18/03/2024	18/03/2024
							F1901	Additional Surgery	A24	No Apparent Device Problem	18/03/2024	18/03/2024
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	18/03/2024	18/03/2024
							F1901	Additional Surgery	A24	No Apparent Device Problem	18/03/2024	18/03/2024
15709	2024	All other reportable incidents	47265	Dual-chamber implantable pacemaker, rate-responsive	CT141	Electrodes/Leads and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT145	Cardiovascular devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT233	Surgical	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT2406	Implantable	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT2407	Active implantable	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT2727	Active implant devices/ancillaries	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT2750	Implantable cardiac electrical stimulators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT2985	Active	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT3001	Implantable metal	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT3002	Implantable synthetic polymer	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT3006	Implantable inorganic material	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT314	Electrically powered	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT321	Long-term surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT449	Generators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT981	Single use	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT983	Surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
15710	2024	All other reportable incidents	37265	Dual-chamber implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A0712	Pacing Problem		
							F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A0712	Pacing Problem		
15711	2024	All other reportable incidents	47205	Cardiac pulse generator programmer	CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A0902	Display or Visual Feedback Problem		
							F27	Problem Identified During Non-Clinical Procedure	A0902	Display or Visual Feedback Problem		
15712	2024	All other reportable incidents	35852	Single-chamber implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E210401	Shock from Patient Lead(s)	A010402	Migration		
									A070909	Oversensing		
									A071302	Inappropriate/Inadequate Shock/Stimulation		
									A1301	Failure to Read Input Signal		
							F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A010402	Migration		
									A070909	Oversensing		
									A071302	Inappropriate/Inadequate Shock/Stimulation		
									A1301	Failure to Read Input Signal		
15713	2024	All other reportable incidents	47265	Dual-chamber implantable pacemaker, rate-responsive	CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	15/08/2024	15/08/2024
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	15/08/2024	15/08/2024
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	15/08/2024	15/08/2024
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	15/08/2024	15/08/2024
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	15/08/2024	15/08/2024

							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	15/08/2024	15/08/2024
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	15/08/2024	15/08/2024
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	15/08/2024	15/08/2024
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	15/08/2024	15/08/2024
15714	2024	All other reportable incidents	47263	Cardiac resynchronization therapy implantable pacemaker	CT1685	Cardiac pacemakers	E2103	Device Overstimulation of Tissue	A0908	Incorrect, Inadequate or Imprecise Result or Readings	16/08/2024	16/08/2024
							F08	Hospitalization or Prolonged Hospitalization	A0908	Incorrect, Inadequate or Imprecise Result or Readings	16/08/2024	16/08/2024
							F1905	Device Revision or Replacement	A0908	Incorrect, Inadequate or Imprecise Result or Readings	16/08/2024	16/08/2024
					CT257	Implantable cardiac pacemakers	E2103	Device Overstimulation of Tissue	A0908	Incorrect, Inadequate or Imprecise Result or Readings	16/08/2024	16/08/2024
							F08	Hospitalization or Prolonged Hospitalization	A0908	Incorrect, Inadequate or Imprecise Result or Readings	16/08/2024	16/08/2024
							F1905	Device Revision or Replacement	A0908	Incorrect, Inadequate or Imprecise Result or Readings	16/08/2024	16/08/2024
					CT2866	Implantable cardiac electrical stimulators	E2103	Device Overstimulation of Tissue	A0908	Incorrect, Inadequate or Imprecise Result or Readings	16/08/2024	16/08/2024
							F08	Hospitalization or Prolonged Hospitalization	A0908	Incorrect, Inadequate or Imprecise Result or Readings	16/08/2024	16/08/2024
							F1905	Device Revision or Replacement	A0908	Incorrect, Inadequate or Imprecise Result or Readings	16/08/2024	16/08/2024
15715	2024	All other reportable incidents	47205	Cardiac pulse generator programmer	CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A0902	Display or Visual Feedback Problem		
							F27	Problem Identified During Non-Clinical Procedure	A0902	Display or Visual Feedback Problem		
15716	2024	All other reportable incidents	47267	Single-chamber implantable pacemaker, rate-responsive	CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
15717	2024	All other reportable incidents	47265	Dual-chamber implantable pacemaker, rate-responsive	CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A0908	Incorrect, Inadequate or Imprecise Result or Readings	02/08/2024	02/08/2024
							F08	Hospitalization or Prolonged Hospitalization	A0908	Incorrect, Inadequate or Imprecise Result or Readings	02/08/2024	02/08/2024
							F1905	Device Revision or Replacement	A0908	Incorrect, Inadequate or Imprecise Result or Readings	02/08/2024	02/08/2024
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A0908	Incorrect, Inadequate or Imprecise Result or Readings	02/08/2024	02/08/2024
							F08	Hospitalization or Prolonged Hospitalization	A0908	Incorrect, Inadequate or Imprecise Result or Readings	02/08/2024	02/08/2024
							F1905	Device Revision or Replacement	A0908	Incorrect, Inadequate or Imprecise Result or Readings	02/08/2024	02/08/2024
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A0908	Incorrect, Inadequate or Imprecise Result or Readings	02/08/2024	02/08/2024
							F08	Hospitalization or Prolonged Hospitalization	A0908	Incorrect, Inadequate or Imprecise Result or Readings	02/08/2024	02/08/2024
							F1905	Device Revision or Replacement	A0908	Incorrect, Inadequate or Imprecise Result or Readings	02/08/2024	02/08/2024
15718	2024	All other reportable incidents	47265	Dual-chamber implantable pacemaker, rate-responsive	CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	18/08/2024	18/08/2024
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	18/08/2024	18/08/2024
							F1903	Device Explantation	A070504	Premature Discharge of Battery	18/08/2024	18/08/2024
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	18/08/2024	18/08/2024
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	18/08/2024	18/08/2024
							F1903	Device Explantation	A070504	Premature Discharge of Battery	18/08/2024	18/08/2024
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	18/08/2024	18/08/2024
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	18/08/2024	18/08/2024
							F1903	Device Explantation	A070504	Premature Discharge of Battery	18/08/2024	18/08/2024
15719	2024	All other reportable incidents	47263	Cardiac resynchronization therapy implantable pacemaker	CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A0908	Incorrect, Inadequate or Imprecise Result or Readings	10/04/2024	10/04/2024
							F08	Hospitalization or Prolonged Hospitalization	A0908	Incorrect, Inadequate or Imprecise Result or Readings	10/04/2024	10/04/2024
							F1905	Device Revision or Replacement	A0908	Incorrect, Inadequate or Imprecise Result or Readings	10/04/2024	10/04/2024
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A0908	Incorrect, Inadequate or Imprecise Result or Readings	10/04/2024	10/04/2024
							F08	Hospitalization or Prolonged Hospitalization	A0908	Incorrect, Inadequate or Imprecise Result or Readings	10/04/2024	10/04/2024
							F1905	Device Revision or Replacement	A0908	Incorrect, Inadequate or Imprecise Result or Readings	10/04/2024	10/04/2024
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A0908	Incorrect, Inadequate or Imprecise Result or Readings	10/04/2024	10/04/2024
							F08	Hospitalization or Prolonged Hospitalization	A0908	Incorrect, Inadequate or Imprecise Result or Readings	10/04/2024	10/04/2024
							F1905	Device Revision or Replacement	A0908	Incorrect, Inadequate or Imprecise Result or Readings	10/04/2024	10/04/2024
15720	2024	All other reportable incidents	47265	Dual-chamber implantable pacemaker, rate-responsive	CT1685	Cardiac pacemakers	E060106	Heart Block	A0712	Pacing Problem		
							F12	Serious Injury/ Illness/ Impairment	A0712	Pacing Problem		
					CT257	Implantable cardiac pacemakers	E060106	Heart Block	A0712	Pacing Problem		
							F12	Serious Injury/ Illness/ Impairment	A0712	Pacing Problem		
					CT2866	Implantable cardiac electrical stimulators	E060106	Heart Block	A0712	Pacing Problem		
							F12	Serious Injury/ Illness/ Impairment	A0712	Pacing Problem		
15721	2024		47265	Dual-chamber implantable pacemaker, rate-responsive	CT1685	Cardiac pacemakers	E011902	Presyncope	A071204	No Pacing		
							E0611	Heart Failure/Congestive Heart Failure	A071204	No Pacing		
							F0101	Therapeutic Response Decreased	A071204	No Pacing		
							F08	Hospitalization or Prolonged Hospitalization	A071204	No Pacing		
							F1905	Device Revision or Replacement	A071204	No Pacing		
					CT257	Implantable cardiac pacemakers	E011902	Presyncope	A071204	No Pacing		
							E0611	Heart Failure/Congestive Heart Failure	A071204	No Pacing		
							F0101	Therapeutic Response Decreased	A071204	No Pacing		
							F08	Hospitalization or Prolonged Hospitalization	A071204	No Pacing		
							F1905	Device Revision or Replacement	A071204	No Pacing		
					CT2866	Implantable cardiac electrical stimulators	E011902	Presyncope	A071204	No Pacing		
							E0611	Heart Failure/Congestive Heart Failure	A071204	No Pacing		
							F0101	Therapeutic Response Decreased	A071204	No Pacing		

							F08	Hospitalization or Prolonged Hospitalization	A071204	No Pacing		
							F1905	Device Revision or Replacement	A071204	No Pacing		
15722	2024	All other reportable incidents	47265	Dual-chamber implantable pacemaker, rate-responsive	CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A0709	Device Sensing Problem	14/08/2024	14/08/2024
									A0712	Pacing Problem	14/08/2024	14/08/2024
							F26	No Health Consequences or Impact	A0709	Device Sensing Problem	14/08/2024	14/08/2024
									A0712	Pacing Problem	14/08/2024	14/08/2024
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A0709	Device Sensing Problem	14/08/2024	14/08/2024
									A0712	Pacing Problem	14/08/2024	14/08/2024
							F26	No Health Consequences or Impact	A0709	Device Sensing Problem	14/08/2024	14/08/2024
									A0712	Pacing Problem	14/08/2024	14/08/2024
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A0709	Device Sensing Problem	14/08/2024	14/08/2024
									A0712	Pacing Problem	14/08/2024	14/08/2024
							F26	No Health Consequences or Impact	A0709	Device Sensing Problem	14/08/2024	14/08/2024
									A0712	Pacing Problem	14/08/2024	14/08/2024
15723	2024	All other reportable incidents	47265	Dual-chamber implantable pacemaker, rate-responsive	CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A070901	Decreased Sensitivity	14/08/2024	14/08/2024
							F26	No Health Consequences or Impact	A070901	Decreased Sensitivity	14/08/2024	14/08/2024
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A070901	Decreased Sensitivity	14/08/2024	14/08/2024
							F26	No Health Consequences or Impact	A070901	Decreased Sensitivity	14/08/2024	14/08/2024
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A070901	Decreased Sensitivity	14/08/2024	14/08/2024
							F26	No Health Consequences or Impact	A070901	Decreased Sensitivity	14/08/2024	14/08/2024
15724	2024	All other reportable incidents	47205	Cardiac pulse generator programmer	CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A0902	Display or Visual Feedback Problem		
							F27	Problem Identified During Non-Clinical Procedure	A0902	Display or Visual Feedback Problem		
15725	2024	All other reportable incidents	47205	Cardiac pulse generator programmer	CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A0902	Display or Visual Feedback Problem		
							F27	Problem Identified During Non-Clinical Procedure	A0902	Display or Visual Feedback Problem		
15726	2024	All other reportable incidents	47265	Dual-chamber implantable pacemaker, rate-responsive	CT1685	Cardiac pacemakers	E060101	Asystole	A0908	Incorrect, Inadequate or Imprecise Result or Readings	19/08/2024	19/08/2024
							F08	Hospitalization or Prolonged Hospitalization	A0908	Incorrect, Inadequate or Imprecise Result or Readings	19/08/2024	19/08/2024
							F1905	Device Revision or Replacement	A0908	Incorrect, Inadequate or Imprecise Result or Readings	19/08/2024	19/08/2024
					CT257	Implantable cardiac pacemakers	E060101	Asystole	A0908	Incorrect, Inadequate or Imprecise Result or Readings	19/08/2024	19/08/2024
							F08	Hospitalization or Prolonged Hospitalization	A0908	Incorrect, Inadequate or Imprecise Result or Readings	19/08/2024	19/08/2024
							F1905	Device Revision or Replacement	A0908	Incorrect, Inadequate or Imprecise Result or Readings	19/08/2024	19/08/2024
					CT2866	Implantable cardiac electrical stimulators	E060101	Asystole	A0908	Incorrect, Inadequate or Imprecise Result or Readings	19/08/2024	19/08/2024
							F08	Hospitalization or Prolonged Hospitalization	A0908	Incorrect, Inadequate or Imprecise Result or Readings	19/08/2024	19/08/2024
							F1905	Device Revision or Replacement	A0908	Incorrect, Inadequate or Imprecise Result or Readings	19/08/2024	19/08/2024
15727	2024	All other reportable incidents	35852	Single-chamber implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E0601	Arrhythmia	A0709	Device Sensing Problem		
									A071201	Failure to Convert Rhythm		
									A071302	Inappropriate/Inadequate Shock/Stimulation		
							F12	Serious Injury/ Illness/ Impairment	A0709	Device Sensing Problem		
									A071201	Failure to Convert Rhythm		
									A071302	Inappropriate/Inadequate Shock/Stimulation		
15728	2024	All other reportable incidents	37265	Dual-chamber implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
									A0908	Incorrect, Inadequate or Imprecise Result or Readings		
							F26	No Health Consequences or Impact	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
									A0908	Incorrect, Inadequate or Imprecise Result or Readings		
15729	2024	All other reportable incidents	47270	Cardiac resynchronization therapy implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem	06/12/2024	06/12/2024
									A070504	Premature Discharge of Battery	06/12/2024	06/12/2024
									A0908	Incorrect, Inadequate or Imprecise Result or Readings	06/12/2024	06/12/2024
							F08	Hospitalization or Prolonged Hospitalization	A0705	Battery Problem	06/12/2024	06/12/2024
									A070504	Premature Discharge of Battery	06/12/2024	06/12/2024
									A0908	Incorrect, Inadequate or Imprecise Result or Readings	06/12/2024	06/12/2024
							F1905	Device Revision or Replacement	A0705	Battery Problem	06/12/2024	06/12/2024
									A070504	Premature Discharge of Battery	06/12/2024	06/12/2024
									A0908	Incorrect, Inadequate or Imprecise Result or Readings	06/12/2024	06/12/2024
15730	2024	All other reportable incidents	47270	Cardiac resynchronization therapy implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem	12/11/2024	12/11/2024
									A070504	Premature Discharge of Battery	12/11/2024	12/11/2024
									A0908	Incorrect, Inadequate or Imprecise Result or Readings	12/11/2024	12/11/2024
							F08	Hospitalization or Prolonged Hospitalization	A0705	Battery Problem	12/11/2024	12/11/2024
									A070504	Premature Discharge of Battery	12/11/2024	12/11/2024
									A0908	Incorrect, Inadequate or Imprecise Result or Readings	12/11/2024	12/11/2024
							F1905	Device Revision or Replacement	A0705	Battery Problem	12/11/2024	12/11/2024
									A070504	Premature Discharge of Battery	12/11/2024	12/11/2024
									A0908	Incorrect, Inadequate or Imprecise Result or Readings	12/11/2024	12/11/2024
15731	2024	All other reportable incidents	17882	Physiologic-monitoring defibrillation system	CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A0719	Unexpected Shutdown		
							F27	Problem Identified During Non-Clinical Procedure	A0719	Unexpected Shutdown		
15732	2024	All other reportable incidents	47265	Dual-chamber implantable pacemaker, rate-responsive	CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A180103	Device Contamination with Body Fluid		
							F12	Serious Injury/ Illness/ Impairment	A180103	Device Contamination with Body Fluid		
							F2307	Reprogramming of Device	A180103	Device Contamination with Body Fluid		
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A180103	Device Contamination with Body Fluid		
							F12	Serious Injury/ Illness/ Impairment	A180103	Device Contamination with Body Fluid		
							F2307	Reprogramming of Device	A180103	Device Contamination with Body Fluid		
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A180103	Device Contamination with Body Fluid		
							F12	Serious Injury/ Illness/ Impairment	A180103	Device Contamination with Body Fluid		
							F2307	Reprogramming of Device	A180103	Device Contamination with Body Fluid		
15733	2024	All other reportable incidents	35852	Single-chamber implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E210401	Shock from Patient Lead(s)	A070504	Premature Discharge of Battery		

									A071302	Inappropriate/Inadequate Shock/Stimulation		
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery		
									A071302	Inappropriate/Inadequate Shock/Stimulation		
							F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A070504	Premature Discharge of Battery		
									A071302	Inappropriate/Inadequate Shock/Stimulation		
15734	2024		47265	Dual-chamber implantable pacemaker, rate-responsive	CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A25	No Apparent Adverse Event	09/12/2023	09/12/2023
							F26	No Health Consequences or Impact	A25	No Apparent Adverse Event	09/12/2023	09/12/2023
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A25	No Apparent Adverse Event	09/12/2023	09/12/2023
							F26	No Health Consequences or Impact	A25	No Apparent Adverse Event	09/12/2023	09/12/2023
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A25	No Apparent Adverse Event	09/12/2023	09/12/2023
							F26	No Health Consequences or Impact	A25	No Apparent Adverse Event	09/12/2023	09/12/2023
15735	2024	All other reportable incidents	17882	Physiologic-monitoring defibrillation system	CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A0719	Unexpected Shutdown		
							F27	Problem Identified During Non-Clinical Procedure	A0719	Unexpected Shutdown		
15736	2024	All other reportable incidents	47265	Dual-chamber implantable pacemaker, rate-responsive	CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
15737	2024	All other reportable incidents	47265	Dual-chamber implantable pacemaker, rate-responsive	CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
15738	2024	All other reportable incidents	47265	Dual-chamber implantable pacemaker, rate-responsive	CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	05/08/2024	05/08/2024
							F1901	Additional Surgery	A24	No Apparent Device Problem	05/08/2024	05/08/2024
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	05/08/2024	05/08/2024
							F1901	Additional Surgery	A24	No Apparent Device Problem	05/08/2024	05/08/2024
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	05/08/2024	05/08/2024
							F1901	Additional Surgery	A24	No Apparent Device Problem	05/08/2024	05/08/2024
15739	2024	All other reportable incidents	37265	Dual-chamber implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E210401	Shock from Patient Lead(s)	A0712	Pacing Problem		
									A071302	Inappropriate/Inadequate Shock/Stimulation		
							F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A0712	Pacing Problem		
									A071302	Inappropriate/Inadequate Shock/Stimulation		
							F1203	Life Threatening Illness or Injury	A0712	Pacing Problem		
									A071302	Inappropriate/Inadequate Shock/Stimulation		
							F2307	Reprogramming of Device	A0712	Pacing Problem		
									A071302	Inappropriate/Inadequate Shock/Stimulation		
15740	2024	All other reportable incidents	47267	Single-chamber implantable pacemaker, rate-responsive	CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A0908	Incorrect, Inadequate or Imprecise Result or Readings	06/08/2024	06/08/2024
							F08	Hospitalization or Prolonged Hospitalization	A0908	Incorrect, Inadequate or Imprecise Result or Readings	06/08/2024	06/08/2024
							F1905	Device Revision or Replacement	A0908	Incorrect, Inadequate or Imprecise Result or Readings	06/08/2024	06/08/2024
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A0908	Incorrect, Inadequate or Imprecise Result or Readings	06/08/2024	06/08/2024
							F08	Hospitalization or Prolonged Hospitalization	A0908	Incorrect, Inadequate or Imprecise Result or Readings	06/08/2024	06/08/2024
							F1905	Device Revision or Replacement	A0908	Incorrect, Inadequate or Imprecise Result or Readings	06/08/2024	06/08/2024
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A0908	Incorrect, Inadequate or Imprecise Result or Readings	06/08/2024	06/08/2024
							F08	Hospitalization or Prolonged Hospitalization	A0908	Incorrect, Inadequate or Imprecise Result or Readings	06/08/2024	06/08/2024
							F1905	Device Revision or Replacement	A0908	Incorrect, Inadequate or Imprecise Result or Readings	06/08/2024	06/08/2024
15741	2024	All other reportable incidents	47267	Single-chamber implantable pacemaker, rate-responsive	CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A0908	Incorrect, Inadequate or Imprecise Result or Readings	06/08/2024	06/08/2024
							F08	Hospitalization or Prolonged Hospitalization	A0908	Incorrect, Inadequate or Imprecise Result or Readings	06/08/2024	06/08/2024
							F1905	Device Revision or Replacement	A0908	Incorrect, Inadequate or Imprecise Result or Readings	06/08/2024	06/08/2024
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A0908	Incorrect, Inadequate or Imprecise Result or Readings	06/08/2024	06/08/2024
							F08	Hospitalization or Prolonged Hospitalization	A0908	Incorrect, Inadequate or Imprecise Result or Readings	06/08/2024	06/08/2024
							F1905	Device Revision or Replacement	A0908	Incorrect, Inadequate or Imprecise Result or Readings	06/08/2024	06/08/2024
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A0908	Incorrect, Inadequate or Imprecise Result or Readings	06/08/2024	06/08/2024
							F08	Hospitalization or Prolonged Hospitalization	A0908	Incorrect, Inadequate or Imprecise Result or Readings	06/08/2024	06/08/2024
							F1905	Device Revision or Replacement	A0908	Incorrect, Inadequate or Imprecise Result or Readings	06/08/2024	06/08/2024
15742	2024	All other reportable incidents	47263	Cardiac resynchronization therapy implantable pacemaker	CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A0908	Incorrect, Inadequate or Imprecise Result or Readings	21/08/2024	21/08/2024
							F08	Hospitalization or Prolonged Hospitalization	A0908	Incorrect, Inadequate or Imprecise Result or Readings	21/08/2024	21/08/2024
							F1905	Device Revision or Replacement	A0908	Incorrect, Inadequate or Imprecise Result or Readings	21/08/2024	21/08/2024

					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A0908	Incorrect, Inadequate or Imprecise Result or Readings	21/08/2024	21/08/2024
							F08	Hospitalization or Prolonged Hospitalization	A0908	Incorrect, Inadequate or Imprecise Result or Readings	21/08/2024	21/08/2024
							F1905	Device Revision or Replacement	A0908	Incorrect, Inadequate or Imprecise Result or Readings	21/08/2024	21/08/2024
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A0908	Incorrect, Inadequate or Imprecise Result or Readings	21/08/2024	21/08/2024
							F08	Hospitalization or Prolonged Hospitalization	A0908	Incorrect, Inadequate or Imprecise Result or Readings	21/08/2024	21/08/2024
							F1905	Device Revision or Replacement	A0908	Incorrect, Inadequate or Imprecise Result or Readings	21/08/2024	21/08/2024
15743	2024		17882	Physiologic-monitoring defibrillation system	CT1684	Cardiac pacemakers and associated devices	E060110	Ventricular Fibrillation	A01	Patient Device Interaction Problem		
									A070603	Failure to Charge		
									A1205	Loose or Intermittent Connection		
							F05	Delay to Treatment/ Therapy	A01	Patient Device Interaction Problem		
									A070603	Failure to Charge		
									A1205	Loose or Intermittent Connection		
15744	2024	All other reportable incidents	47265	Dual-chamber implantable pacemaker, rate-responsive	CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
15745	2024	All other reportable incidents	17882	Physiologic-monitoring defibrillation system	CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070802	Intermittent Loss of Power		
							F27	Problem Identified During Non-Clinical Procedure	A070802	Intermittent Loss of Power		
15746	2024	All other reportable incidents	17882	Physiologic-monitoring defibrillation system	CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070802	Intermittent Loss of Power		
							F27	Problem Identified During Non-Clinical Procedure	A070802	Intermittent Loss of Power		
15747	2024	All other reportable incidents	47265	Dual-chamber implantable pacemaker, rate-responsive	CT1685	Cardiac pacemakers	E0717	Dyspnea	A071102	Failure to Interrogate	21/08/2024	21/08/2024
									A071204	No Pacing	21/08/2024	21/08/2024
							F1901	Additional Surgery	A071102	Failure to Interrogate	21/08/2024	21/08/2024
									A071204	No Pacing	21/08/2024	21/08/2024
					CT257	Implantable cardiac pacemakers	E0717	Dyspnea	A071102	Failure to Interrogate	21/08/2024	21/08/2024
									A071204	No Pacing	21/08/2024	21/08/2024
							F1901	Additional Surgery	A071102	Failure to Interrogate	21/08/2024	21/08/2024
									A071204	No Pacing	21/08/2024	21/08/2024
					CT2866	Implantable cardiac electrical stimulators	E0717	Dyspnea	A071102	Failure to Interrogate	21/08/2024	21/08/2024
									A071204	No Pacing	21/08/2024	21/08/2024
							F1901	Additional Surgery	A071102	Failure to Interrogate	21/08/2024	21/08/2024
									A071204	No Pacing	21/08/2024	21/08/2024
15748	2024	All other reportable incidents	47265	Dual-chamber implantable pacemaker, rate-responsive	CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A0908	Incorrect, Inadequate or Imprecise Result or Readings	02/02/2024	02/02/2024
							F08	Hospitalization or Prolonged Hospitalization	A0908	Incorrect, Inadequate or Imprecise Result or Readings	02/02/2024	02/02/2024
							F1905	Device Revision or Replacement	A0908	Incorrect, Inadequate or Imprecise Result or Readings	02/02/2024	02/02/2024
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A0908	Incorrect, Inadequate or Imprecise Result or Readings	02/02/2024	02/02/2024
							F08	Hospitalization or Prolonged Hospitalization	A0908	Incorrect, Inadequate or Imprecise Result or Readings	02/02/2024	02/02/2024
							F1905	Device Revision or Replacement	A0908	Incorrect, Inadequate or Imprecise Result or Readings	02/02/2024	02/02/2024
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A0908	Incorrect, Inadequate or Imprecise Result or Readings	02/02/2024	02/02/2024
							F08	Hospitalization or Prolonged Hospitalization	A0908	Incorrect, Inadequate or Imprecise Result or Readings	02/02/2024	02/02/2024
							F1905	Device Revision or Replacement	A0908	Incorrect, Inadequate or Imprecise Result or Readings	02/02/2024	02/02/2024
15749	2024	All other reportable incidents	47263	Cardiac resynchronization therapy implantable pacemaker	CT1685	Cardiac pacemakers	E011902	Presyncope	A070909	Oversensing	11/04/2024	11/04/2024
									A0712	Pacing Problem	11/04/2024	11/04/2024
									A0908	Incorrect, Inadequate or Imprecise Result or Readings	11/04/2024	11/04/2024
							F08	Hospitalization or Prolonged Hospitalization	A070909	Oversensing	11/04/2024	11/04/2024
									A0712	Pacing Problem	11/04/2024	11/04/2024
									A0908	Incorrect, Inadequate or Imprecise Result or Readings	11/04/2024	11/04/2024
							F0801	Intensive Care	A070909	Oversensing	11/04/2024	11/04/2024
									A0712	Pacing Problem	11/04/2024	11/04/2024
									A0908	Incorrect, Inadequate or Imprecise Result or Readings	11/04/2024	11/04/2024
							F1001	Absence of Treatment	A070909	Oversensing	11/04/2024	11/04/2024
									A0712	Pacing Problem	11/04/2024	11/04/2024
									A0908	Incorrect, Inadequate or Imprecise Result or Readings	11/04/2024	11/04/2024
							F1905	Device Revision or Replacement	A070909	Oversensing	11/04/2024	11/04/2024
									A0712	Pacing Problem	11/04/2024	11/04/2024
									A0908	Incorrect, Inadequate or Imprecise Result or Readings	11/04/2024	11/04/2024
					CT257	Implantable cardiac pacemakers	E011902	Presyncope	A070909	Oversensing	11/04/2024	11/04/2024
									A0712	Pacing Problem	11/04/2024	11/04/2024
									A0908	Incorrect, Inadequate or Imprecise Result or Readings	11/04/2024	11/04/2024
							F08	Hospitalization or Prolonged Hospitalization	A070909	Oversensing	11/04/2024	11/04/2024
									A0712	Pacing Problem	11/04/2024	11/04/2024
									A0908	Incorrect, Inadequate or Imprecise Result or Readings	11/04/2024	11/04/2024
							F0801	Intensive Care	A070909	Oversensing	11/04/2024	11/04/2024
									A0712	Pacing Problem	11/04/2024	11/04/2024
									A0908	Incorrect, Inadequate or Imprecise Result or Readings	11/04/2024	11/04/2024
							F1001	Absence of Treatment	A070909	Oversensing	11/04/2024	11/04/2024
									A0712	Pacing Problem	11/04/2024	11/04/2024

									A0908	Incorrect, Inadequate or Imprecise Result or Readings	11/04/2024	11/04/2024
							F1905	Device Revision or Replacement	A070909	Oversensing	11/04/2024	11/04/2024
									A0712	Pacing Problem	11/04/2024	11/04/2024
									A0908	Incorrect, Inadequate or Imprecise Result or Readings	11/04/2024	11/04/2024
					CT2866	Implantable cardiac electrical stimulators	E011902	Presyncope	A070909	Oversensing	11/04/2024	11/04/2024
									A0712	Pacing Problem	11/04/2024	11/04/2024
									A0908	Incorrect, Inadequate or Imprecise Result or Readings	11/04/2024	11/04/2024
							F08	Hospitalization or Prolonged Hospitalization	A070909	Oversensing	11/04/2024	11/04/2024
									A0712	Pacing Problem	11/04/2024	11/04/2024
									A0908	Incorrect, Inadequate or Imprecise Result or Readings	11/04/2024	11/04/2024
							F0801	Intensive Care	A070909	Oversensing	11/04/2024	11/04/2024
									A0712	Pacing Problem	11/04/2024	11/04/2024
									A0908	Incorrect, Inadequate or Imprecise Result or Readings	11/04/2024	11/04/2024
							F1001	Absence of Treatment	A070909	Oversensing	11/04/2024	11/04/2024
									A0712	Pacing Problem	11/04/2024	11/04/2024
									A0908	Incorrect, Inadequate or Imprecise Result or Readings	11/04/2024	11/04/2024
							F1905	Device Revision or Replacement	A070909	Oversensing	11/04/2024	11/04/2024
									A0712	Pacing Problem	11/04/2024	11/04/2024
									A0908	Incorrect, Inadequate or Imprecise Result or Readings	11/04/2024	11/04/2024
15750	2024	All other reportable incidents	47265	Dual-chamber implantable pacemaker, rate-responsive	CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	21/08/2024	21/08/2024
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	21/08/2024	21/08/2024
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	21/08/2024	21/08/2024
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	21/08/2024	21/08/2024
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	21/08/2024	21/08/2024
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	21/08/2024	21/08/2024
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	21/08/2024	21/08/2024
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	21/08/2024	21/08/2024
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	21/08/2024	21/08/2024
15751	2024	All other reportable incidents	35852	Single-chamber implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	08/07/2025	08/07/2025
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	08/07/2025	08/07/2025
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	08/07/2025	08/07/2025
15752	2024		47265	Dual-chamber implantable pacemaker, rate-responsive	CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	17/08/2024	17/08/2024
							F1901	Additional Surgery	A24	No Apparent Device Problem	17/08/2024	17/08/2024
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	17/08/2024	17/08/2024
							F1901	Additional Surgery	A24	No Apparent Device Problem	17/08/2024	17/08/2024
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	17/08/2024	17/08/2024
							F1901	Additional Surgery	A24	No Apparent Device Problem	17/08/2024	17/08/2024
15753	2024	All other reportable incidents	35852	Single-chamber implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
									A0908	Incorrect, Inadequate or Imprecise Result or Readings		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
									A0908	Incorrect, Inadequate or Imprecise Result or Readings		
15754	2024	All other reportable incidents	37265	Dual-chamber implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
15755	2024	All other reportable incidents	47205	Cardiac pulse generator programmer	CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A09	Output Problem		
									A13	Communication or Transmission Problem		
							F26	No Health Consequences or Impact	A09	Output Problem		
									A13	Communication or Transmission Problem		
15756	2024	All other reportable incidents	36106	Implantable pacemaker analyser, intraoperative	CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A09	Output Problem		
									A0902	Display or Visual Feedback Problem		
									A13	Communication or Transmission Problem		
							F23	Unexpected Medical Intervention	A09	Output Problem		
									A0902	Display or Visual Feedback Problem		
									A13	Communication or Transmission Problem		
15757	2024	All other reportable incidents	47270	Cardiac resynchronization therapy implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A070909	Oversensing		
									A0712	Pacing Problem		
									A071204	No Pacing		
							F23	Unexpected Medical Intervention	A070909	Oversensing		
									A0712	Pacing Problem		
									A071204	No Pacing		
15758	2024	All other reportable incidents			CT1684	Cardiac pacemakers and associated devices	E0601	Arrhythmia	A070909	Oversensing		
							F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A070909	Oversensing		
							F15	Recognized Device or Procedural Complication	A070909	Oversensing		
15759	2024	All other reportable incidents	47265	Dual-chamber implantable pacemaker, rate-responsive	CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F1901	Additional Surgery	A070504	Premature Discharge of Battery		
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F1901	Additional Surgery	A070504	Premature Discharge of Battery		
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F1901	Additional Surgery	A070504	Premature Discharge of Battery		
15760	2024	All other reportable incidents	47265	Dual-chamber implantable pacemaker, rate-responsive	CT1685	Cardiac pacemakers	E0112	Dizziness	A070101	Failure to Capture	22/01/2024	22/01/2024
									A0908	Incorrect, Inadequate or Imprecise Result or Readings	22/01/2024	22/01/2024
							E060101	Asystole	A070101	Failure to Capture	22/01/2024	22/01/2024
									A0908	Incorrect, Inadequate or Imprecise Result or Readings	22/01/2024	22/01/2024
							F08	Hospitalization or Prolonged Hospitalization	A070101	Failure to Capture	22/01/2024	22/01/2024
									A0908	Incorrect, Inadequate or Imprecise Result or Readings	22/01/2024	22/01/2024
							F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A070101	Failure to Capture	22/01/2024	22/01/2024
									A0908	Incorrect, Inadequate or Imprecise Result or Readings	22/01/2024	22/01/2024
							F1905	Device Revision or Replacement	A070101	Failure to Capture	22/01/2024	22/01/2024

									A0908	Incorrect, Inadequate or Imprecise Result or Readings	22/01/2024	22/01/2024
					CT257	Implantable cardiac pacemakers	E0112	Dizziness	A070101	Failure to Capture	22/01/2024	22/01/2024
									A0908	Incorrect, Inadequate or Imprecise Result or Readings	22/01/2024	22/01/2024
							E060101	Asystole	A070101	Failure to Capture	22/01/2024	22/01/2024
									A0908	Incorrect, Inadequate or Imprecise Result or Readings	22/01/2024	22/01/2024
							F08	Hospitalization or Prolonged Hospitalization	A070101	Failure to Capture	22/01/2024	22/01/2024
									A0908	Incorrect, Inadequate or Imprecise Result or Readings	22/01/2024	22/01/2024
							F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A070101	Failure to Capture	22/01/2024	22/01/2024
									A0908	Incorrect, Inadequate or Imprecise Result or Readings	22/01/2024	22/01/2024
							F1905	Device Revision or Replacement	A070101	Failure to Capture	22/01/2024	22/01/2024
									A0908	Incorrect, Inadequate or Imprecise Result or Readings	22/01/2024	22/01/2024
					CT2866	Implantable cardiac electrical stimulators	E0112	Dizziness	A070101	Failure to Capture	22/01/2024	22/01/2024
									A0908	Incorrect, Inadequate or Imprecise Result or Readings	22/01/2024	22/01/2024
							E060101	Asystole	A070101	Failure to Capture	22/01/2024	22/01/2024
									A0908	Incorrect, Inadequate or Imprecise Result or Readings	22/01/2024	22/01/2024
							F08	Hospitalization or Prolonged Hospitalization	A070101	Failure to Capture	22/01/2024	22/01/2024
									A0908	Incorrect, Inadequate or Imprecise Result or Readings	22/01/2024	22/01/2024
							F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A070101	Failure to Capture	22/01/2024	22/01/2024
									A0908	Incorrect, Inadequate or Imprecise Result or Readings	22/01/2024	22/01/2024
							F1905	Device Revision or Replacement	A070101	Failure to Capture	22/01/2024	22/01/2024
									A0908	Incorrect, Inadequate or Imprecise Result or Readings	22/01/2024	22/01/2024
15761	2024	All other reportable incidents	47265	Dual-chamber implantable pacemaker, rate-responsive	CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem	22/04/2024	22/04/2024
									A070504	Premature Discharge of Battery	22/04/2024	22/04/2024
							F08	Hospitalization or Prolonged Hospitalization	A0705	Battery Problem	22/04/2024	22/04/2024
									A070504	Premature Discharge of Battery	22/04/2024	22/04/2024
							F1905	Device Revision or Replacement	A0705	Battery Problem	22/04/2024	22/04/2024
									A070504	Premature Discharge of Battery	22/04/2024	22/04/2024
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem	22/04/2024	22/04/2024
									A070504	Premature Discharge of Battery	22/04/2024	22/04/2024
							F08	Hospitalization or Prolonged Hospitalization	A0705	Battery Problem	22/04/2024	22/04/2024
									A070504	Premature Discharge of Battery	22/04/2024	22/04/2024
							F1905	Device Revision or Replacement	A0705	Battery Problem	22/04/2024	22/04/2024
									A070504	Premature Discharge of Battery	22/04/2024	22/04/2024
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem	22/04/2024	22/04/2024
									A070504	Premature Discharge of Battery	22/04/2024	22/04/2024
							F08	Hospitalization or Prolonged Hospitalization	A0705	Battery Problem	22/04/2024	22/04/2024
									A070504	Premature Discharge of Battery	22/04/2024	22/04/2024
							F1905	Device Revision or Replacement	A0705	Battery Problem	22/04/2024	22/04/2024
									A070504	Premature Discharge of Battery	22/04/2024	22/04/2024
15762	2024	All other reportable incidents	47265	Dual-chamber implantable pacemaker, rate-responsive								
15763	2024	All other reportable incidents	35852	Single-chamber implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E210401	Shock from Patient Lead(s)	A0709	Device Sensing Problem		
									A071302	Inappropriate/Inadequate Shock/Stimulation		
							F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A0709	Device Sensing Problem		
									A071302	Inappropriate/Inadequate Shock/Stimulation		
							F1203	Life Threatening Illness or Injury	A0709	Device Sensing Problem		
									A071302	Inappropriate/Inadequate Shock/Stimulation		
15764	2024	All other reportable incidents	35852	Single-chamber implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	02/07/2024	02/07/2024
									A071102	Failure to Interrogate	02/07/2024	02/07/2024
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	02/07/2024	02/07/2024
									A071102	Failure to Interrogate	02/07/2024	02/07/2024
							F12	Serious Injury/ Illness/ Impairment	A070504	Premature Discharge of Battery	02/07/2024	02/07/2024
									A071102	Failure to Interrogate	02/07/2024	02/07/2024
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	02/07/2024	02/07/2024
									A071102	Failure to Interrogate	02/07/2024	02/07/2024
15765	2024	All other reportable incidents	36106	Implantable pacemaker analyser, intraoperative	CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A0902	Display or Visual Feedback Problem		
									A13	Communication or Transmission Problem		
							F26	No Health Consequences or Impact	A0902	Display or Visual Feedback Problem		
									A13	Communication or Transmission Problem		
15766	2024	All other reportable incidents	35852	Single-chamber implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A070909	Oversensing		
									A090801	Signal Artifact/Noise		
							F12	Serious Injury/ Illness/ Impairment	A070909	Oversensing		
									A090801	Signal Artifact/Noise		
							F2307	Reprogramming of Device	A070909	Oversensing		
									A090801	Signal Artifact/Noise		
15767	2024	All other reportable incidents	36106	Implantable pacemaker analyser, intraoperative	CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A0902	Display or Visual Feedback Problem		
									A11	Computer Software Problem		
							F27	Problem Identified During Non-Clinical Procedure	A0902	Display or Visual Feedback Problem		
									A11	Computer Software Problem		
15768	2024	All other reportable incidents	17882	Physiologic-monitoring defibrillation system	CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A0719	Unexpected Shutdown		
							F27	Problem Identified During Non-Clinical Procedure	A0719	Unexpected Shutdown		
15769	2024	All other reportable incidents	37265	Dual-chamber implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E210401	Shock from Patient Lead(s)	A0709	Device Sensing Problem		
									A0712	Pacing Problem		
									A071302	Inappropriate/Inadequate Shock/Stimulation		
							F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A0709	Device Sensing Problem		
									A0712	Pacing Problem		
									A071302	Inappropriate/Inadequate Shock/Stimulation		
							F1203	Life Threatening Illness or Injury	A0709	Device Sensing Problem		
									A0712	Pacing Problem		
									A071302	Inappropriate/Inadequate Shock/Stimulation		

15770	2024	All other reportable incidents	47265	Dual-chamber implantable pacemaker, rate-responsive	CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A070101	Failure to Capture		
							F26	No Health Consequences or Impact	A070101	Failure to Capture		
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A070101	Failure to Capture		
							F26	No Health Consequences or Impact	A070101	Failure to Capture		
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A070101	Failure to Capture		
							F26	No Health Consequences or Impact	A070101	Failure to Capture		
15771	2024	All other reportable incidents	35852	Single-chamber implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem	03/11/2023	03/11/2023
									A070504	Premature Discharge of Battery	03/11/2023	03/11/2023
							F08	Hospitalization or Prolonged Hospitalization	A0705	Battery Problem	03/11/2023	03/11/2023
									A070504	Premature Discharge of Battery	03/11/2023	03/11/2023
							F1903	Device Explantation	A0705	Battery Problem	03/11/2023	03/11/2023
									A070504	Premature Discharge of Battery	03/11/2023	03/11/2023
15772	2024	All other reportable incidents	47263	Cardiac resynchronization therapy implantable pacemaker	CT1685	Cardiac pacemakers	E011903	Syncope/Fainting	A070101	Failure to Capture	04/10/2023	04/10/2023
									A0908	Incorrect, Inadequate or Imprecise Result or Readings	04/10/2023	04/10/2023
							E060104	Bradycardia	A070101	Failure to Capture	04/10/2023	04/10/2023
									A0908	Incorrect, Inadequate or Imprecise Result or Readings	04/10/2023	04/10/2023
							F08	Hospitalization or Prolonged Hospitalization	A070101	Failure to Capture	04/10/2023	04/10/2023
									A0908	Incorrect, Inadequate or Imprecise Result or Readings	04/10/2023	04/10/2023
							F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A070101	Failure to Capture	04/10/2023	04/10/2023
									A0908	Incorrect, Inadequate or Imprecise Result or Readings	04/10/2023	04/10/2023
							F1905	Device Revision or Replacement	A070101	Failure to Capture	04/10/2023	04/10/2023
									A0908	Incorrect, Inadequate or Imprecise Result or Readings	04/10/2023	04/10/2023
					CT257	Implantable cardiac pacemakers	E011903	Syncope/Fainting	A070101	Failure to Capture	04/10/2023	04/10/2023
									A0908	Incorrect, Inadequate or Imprecise Result or Readings	04/10/2023	04/10/2023
							E060104	Bradycardia	A070101	Failure to Capture	04/10/2023	04/10/2023
									A0908	Incorrect, Inadequate or Imprecise Result or Readings	04/10/2023	04/10/2023
							F08	Hospitalization or Prolonged Hospitalization	A070101	Failure to Capture	04/10/2023	04/10/2023
									A0908	Incorrect, Inadequate or Imprecise Result or Readings	04/10/2023	04/10/2023
							F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A070101	Failure to Capture	04/10/2023	04/10/2023
									A0908	Incorrect, Inadequate or Imprecise Result or Readings	04/10/2023	04/10/2023
							F1905	Device Revision or Replacement	A070101	Failure to Capture	04/10/2023	04/10/2023
									A0908	Incorrect, Inadequate or Imprecise Result or Readings	04/10/2023	04/10/2023
					CT2866	Implantable cardiac electrical stimulators	E011903	Syncope/Fainting	A070101	Failure to Capture	04/10/2023	04/10/2023
									A0908	Incorrect, Inadequate or Imprecise Result or Readings	04/10/2023	04/10/2023
							E060104	Bradycardia	A070101	Failure to Capture	04/10/2023	04/10/2023
									A0908	Incorrect, Inadequate or Imprecise Result or Readings	04/10/2023	04/10/2023
							F08	Hospitalization or Prolonged Hospitalization	A070101	Failure to Capture	04/10/2023	04/10/2023
									A0908	Incorrect, Inadequate or Imprecise Result or Readings	04/10/2023	04/10/2023
							F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A070101	Failure to Capture	04/10/2023	04/10/2023
									A0908	Incorrect, Inadequate or Imprecise Result or Readings	04/10/2023	04/10/2023
							F1905	Device Revision or Replacement	A070101	Failure to Capture	04/10/2023	04/10/2023
									A0908	Incorrect, Inadequate or Imprecise Result or Readings	04/10/2023	04/10/2023
15773	2024	All other reportable incidents	47265	Dual-chamber implantable pacemaker, rate-responsive	CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A0908	Incorrect, Inadequate or Imprecise Result or Readings	23/08/2024	23/08/2024
							F08	Hospitalization or Prolonged Hospitalization	A0908	Incorrect, Inadequate or Imprecise Result or Readings	23/08/2024	23/08/2024
							F1905	Device Revision or Replacement	A0908	Incorrect, Inadequate or Imprecise Result or Readings	23/08/2024	23/08/2024
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A0908	Incorrect, Inadequate or Imprecise Result or Readings	23/08/2024	23/08/2024
							F08	Hospitalization or Prolonged Hospitalization	A0908	Incorrect, Inadequate or Imprecise Result or Readings	23/08/2024	23/08/2024
							F1905	Device Revision or Replacement	A0908	Incorrect, Inadequate or Imprecise Result or Readings	23/08/2024	23/08/2024
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A0908	Incorrect, Inadequate or Imprecise Result or Readings	23/08/2024	23/08/2024
							F08	Hospitalization or Prolonged Hospitalization	A0908	Incorrect, Inadequate or Imprecise Result or Readings	23/08/2024	23/08/2024
							F1905	Device Revision or Replacement	A0908	Incorrect, Inadequate or Imprecise Result or Readings	23/08/2024	23/08/2024
15774	2024	All other reportable incidents	17882	Physiologic-monitoring defibrillation system	CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070803	Failure to Power Up		
							F27	Problem Identified During Non-Clinical Procedure	A070803	Failure to Power Up		
15775	2024	All other reportable incidents	17882	Physiologic-monitoring defibrillation system	CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A0719	Unexpected Shutdown		
							F26	No Health Consequences or Impact	A0719	Unexpected Shutdown		
15776	2024	All other reportable incidents	17882	Physiologic-monitoring defibrillation system	CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A11	Computer Software Problem		
							F27	Problem Identified During Non-Clinical Procedure	A11	Computer Software Problem		
15777	2024	All other reportable incidents	17882	Physiologic-monitoring defibrillation system	CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A0719	Unexpected Shutdown		
							F27	Problem Identified During Non-Clinical Procedure	A0719	Unexpected Shutdown		
15778	2024	All other reportable incidents	17882	Physiologic-monitoring defibrillation system	CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A160503	Inappropriate or Unexpected Reset		
							F27	Problem Identified During Non-Clinical Procedure	A160503	Inappropriate or Unexpected Reset		
15779	2024	All other reportable incidents	36106	Implantable pacemaker analyser, intraoperative	CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A1104	Computer Operating System Problem		
							F26	No Health Consequences or Impact	A1104	Computer Operating System Problem		
15780	2024	All other reportable incidents	47265	Dual-chamber implantable pacemaker, rate-responsive	CT1685	Cardiac pacemakers	E060109	Tachycardia	A090811	Unable to Obtain Readings		
							E2403	No Clinical Signs, Symptoms or Conditions	A090811	Unable to Obtain Readings		
							F11	Non-Serious Injury/ Illness/ Impairment	A090811	Unable to Obtain Readings		
							F26	No Health Consequences or Impact	A090811	Unable to Obtain Readings		
					CT257	Implantable cardiac pacemakers	E060109	Tachycardia	A090811	Unable to Obtain Readings		
							E2403	No Clinical Signs, Symptoms or Conditions	A090811	Unable to Obtain Readings		

							F11	Non-Serious Injury/ Illness/ Impairment	A090811	Unable to Obtain Readings		
							F26	No Health Consequences or Impact	A090811	Unable to Obtain Readings		
					CT2866	Implantable cardiac electrical stimulators	E060109	Tachycardia	A090811	Unable to Obtain Readings		
							E2403	No Clinical Signs, Symptoms or Conditions	A090811	Unable to Obtain Readings		
							F11	Non-Serious Injury/ Illness/ Impairment	A090811	Unable to Obtain Readings		
							F26	No Health Consequences or Impact	A090811	Unable to Obtain Readings		
15781	2024	All other reportable incidents	47265	Dual-chamber implantable pacemaker, rate-responsive	CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A070909	Oversensing		
									A072201	High Impedance		
									A090801	Signal Artifact/Noise		
							F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A070909	Oversensing		
									A072201	High Impedance		
									A090801	Signal Artifact/Noise		
							F2203	Imaging Required	A070909	Oversensing		
									A072201	High Impedance		
									A090801	Signal Artifact/Noise		
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A070909	Oversensing		
									A072201	High Impedance		
									A090801	Signal Artifact/Noise		
							F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A070909	Oversensing		
									A072201	High Impedance		
									A090801	Signal Artifact/Noise		
							F2203	Imaging Required	A070909	Oversensing		
									A072201	High Impedance		
									A090801	Signal Artifact/Noise		
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A070909	Oversensing		
									A072201	High Impedance		
									A090801	Signal Artifact/Noise		
							F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A070909	Oversensing		
									A072201	High Impedance		
									A090801	Signal Artifact/Noise		
							F2203	Imaging Required	A070909	Oversensing		
									A072201	High Impedance		
									A090801	Signal Artifact/Noise		
15782	2024	All other reportable incidents	47265	Dual-chamber implantable pacemaker, rate-responsive	CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A05	Mechanical Problem		
									A1205	Loose or Intermittent Connection		
							F26	No Health Consequences or Impact	A05	Mechanical Problem		
									A1205	Loose or Intermittent Connection		
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A05	Mechanical Problem		
									A1205	Loose or Intermittent Connection		
							F26	No Health Consequences or Impact	A05	Mechanical Problem		
									A1205	Loose or Intermittent Connection		
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A05	Mechanical Problem		
									A1205	Loose or Intermittent Connection		
							F26	No Health Consequences or Impact	A05	Mechanical Problem		
									A1205	Loose or Intermittent Connection		
15783	2024	All other reportable incidents	47265	Dual-chamber implantable pacemaker, rate-responsive	CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	01/12/2023	01/12/2023
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	01/12/2023	01/12/2023
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	01/12/2023	01/12/2023
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	01/12/2023	01/12/2023
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	01/12/2023	01/12/2023
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	01/12/2023	01/12/2023
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	01/12/2023	01/12/2023
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	01/12/2023	01/12/2023
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	01/12/2023	01/12/2023
15784	2024	All other reportable incidents	37265	Dual-chamber implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A0712	Pacing Problem	03/06/2024	03/06/2024
									A1206	Misconnection	03/06/2024	03/06/2024
							F12	Serious Injury/ Illness/ Impairment	A0712	Pacing Problem	03/06/2024	03/06/2024
									A1206	Misconnection	03/06/2024	03/06/2024
							F23	Unexpected Medical Intervention	A0712	Pacing Problem	03/06/2024	03/06/2024
									A1206	Misconnection	03/06/2024	03/06/2024
15785	2024	All other reportable incidents	47265	Dual-chamber implantable pacemaker, rate-responsive	CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A070101	Failure to Capture		
									A070910	Undersensing		
							F26	No Health Consequences or Impact	A070101	Failure to Capture		
									A070910	Undersensing		
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A070101	Failure to Capture		
									A070910	Undersensing		
							F26	No Health Consequences or Impact	A070101	Failure to Capture		
									A070910	Undersensing		
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A070101	Failure to Capture		
									A070910	Undersensing		
							F26	No Health Consequences or Impact	A070101	Failure to Capture		
									A070910	Undersensing		
15786	2024	All other reportable incidents	47265	Dual-chamber implantable pacemaker, rate-responsive	CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	22/12/2023	22/12/2023
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	22/12/2023	22/12/2023
							F1903	Device Explantation	A070504	Premature Discharge of Battery	22/12/2023	22/12/2023
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	22/12/2023	22/12/2023
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	22/12/2023	22/12/2023
							F1903	Device Explantation	A070504	Premature Discharge of Battery	22/12/2023	22/12/2023
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	22/12/2023	22/12/2023
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	22/12/2023	22/12/2023
							F1903	Device Explantation	A070504	Premature Discharge of Battery	22/12/2023	22/12/2023
15787	2024	All other reportable incidents	17882	Physiologic-monitoring defibrillation system	CT1684	Cardiac pacemakers and associated devices	E060110	Ventricular Fibrillation	A01	Patient Device Interaction Problem		
									A070603	Failure to Charge		
									A1205	Loose or Intermittent Connection		
							F05	Delay to Treatment/ Therapy	A01	Patient Device Interaction Problem		
									A070603	Failure to Charge		
									A1205	Loose or Intermittent Connection		
15788	2024	All other reportable incidents	47270	Cardiac resynchronization therapy implantable	CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A0203	Defective Device		

				defibrillator					A070909	Oversensing		
									A070910	Undersensing		
									A072202	Low Impedance		
									A090801	Signal Artifact/Noise		
									A1301	Failure to Read Input Signal		
							F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A0203	Defective Device		
									A070909	Oversensing		
									A070910	Undersensing		
									A072202	Low Impedance		
									A090801	Signal Artifact/Noise		
									A1301	Failure to Read Input Signal		
15789	2024	All other reportable incidents	47265	Dual-chamber implantable pacemaker, rate-responsive	CT1685	Cardiac pacemakers	E0601	Arrhythmia	A070504	Premature Discharge of Battery	06/06/2024	06/06/2024
									A071102	Failure to Interrogate	06/06/2024	06/06/2024
									A071204	No Pacing	06/06/2024	06/06/2024
							F12	Serious Injury/ Illness/ Impairment	A070504	Premature Discharge of Battery	06/06/2024	06/06/2024
									A071102	Failure to Interrogate	06/06/2024	06/06/2024
									A071204	No Pacing	06/06/2024	06/06/2024
					CT257	Implantable cardiac pacemakers	E0601	Arrhythmia	A070504	Premature Discharge of Battery	06/06/2024	06/06/2024
									A071102	Failure to Interrogate	06/06/2024	06/06/2024
									A071204	No Pacing	06/06/2024	06/06/2024
							F12	Serious Injury/ Illness/ Impairment	A070504	Premature Discharge of Battery	06/06/2024	06/06/2024
									A071102	Failure to Interrogate	06/06/2024	06/06/2024
									A071204	No Pacing	06/06/2024	06/06/2024
					CT2866	Implantable cardiac electrical stimulators	E0601	Arrhythmia	A070504	Premature Discharge of Battery	06/06/2024	06/06/2024
									A071102	Failure to Interrogate	06/06/2024	06/06/2024
									A071204	No Pacing	06/06/2024	06/06/2024
							F12	Serious Injury/ Illness/ Impairment	A070504	Premature Discharge of Battery	06/06/2024	06/06/2024
									A071102	Failure to Interrogate	06/06/2024	06/06/2024
									A071204	No Pacing	06/06/2024	06/06/2024
15790	2024	All other reportable incidents	47265	Dual-chamber implantable pacemaker, rate-responsive	CT141	Electrodes/Leads and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A071102	Failure to Interrogate	06/06/2024	06/06/2024
									A071204	No Pacing	06/06/2024	06/06/2024
							F1901	Additional Surgery	A071102	Failure to Interrogate	06/06/2024	06/06/2024
									A071204	No Pacing	06/06/2024	06/06/2024
					CT145	Cardiovascular devices	E2403	No Clinical Signs, Symptoms or Conditions	A071102	Failure to Interrogate	06/06/2024	06/06/2024
									A071204	No Pacing	06/06/2024	06/06/2024
							F1901	Additional Surgery	A071102	Failure to Interrogate	06/06/2024	06/06/2024
									A071204	No Pacing	06/06/2024	06/06/2024
					CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A071102	Failure to Interrogate	06/06/2024	06/06/2024
									A071204	No Pacing	06/06/2024	06/06/2024
							F1901	Additional Surgery	A071102	Failure to Interrogate	06/06/2024	06/06/2024
									A071204	No Pacing	06/06/2024	06/06/2024
					CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A071102	Failure to Interrogate	06/06/2024	06/06/2024
									A071204	No Pacing	06/06/2024	06/06/2024
							F1901	Additional Surgery	A071102	Failure to Interrogate	06/06/2024	06/06/2024
									A071204	No Pacing	06/06/2024	06/06/2024
					CT233	Surgical	E2403	No Clinical Signs, Symptoms or Conditions	A071102	Failure to Interrogate	06/06/2024	06/06/2024
									A071204	No Pacing	06/06/2024	06/06/2024
							F1901	Additional Surgery	A071102	Failure to Interrogate	06/06/2024	06/06/2024
									A071204	No Pacing	06/06/2024	06/06/2024
					CT2406	Implantable	E2403	No Clinical Signs, Symptoms or Conditions	A071102	Failure to Interrogate	06/06/2024	06/06/2024
									A071204	No Pacing	06/06/2024	06/06/2024
							F1901	Additional Surgery	A071102	Failure to Interrogate	06/06/2024	06/06/2024
									A071204	No Pacing	06/06/2024	06/06/2024
					CT2407	Active implantable	E2403	No Clinical Signs, Symptoms or Conditions	A071102	Failure to Interrogate	06/06/2024	06/06/2024
									A071204	No Pacing	06/06/2024	06/06/2024
							F1901	Additional Surgery	A071102	Failure to Interrogate	06/06/2024	06/06/2024
									A071204	No Pacing	06/06/2024	06/06/2024
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A071102	Failure to Interrogate	06/06/2024	06/06/2024
									A071204	No Pacing	06/06/2024	06/06/2024
							F1901	Additional Surgery	A071102	Failure to Interrogate	06/06/2024	06/06/2024
									A071204	No Pacing	06/06/2024	06/06/2024
					CT2727	Active implant devices/ancillaries	E2403	No Clinical Signs, Symptoms or Conditions	A071102	Failure to Interrogate	06/06/2024	06/06/2024
									A071204	No Pacing	06/06/2024	06/06/2024
							F1901	Additional Surgery	A071102	Failure to Interrogate	06/06/2024	06/06/2024
									A071204	No Pacing	06/06/2024	06/06/2024
					CT2750	Implantable cardiac electrical stimulators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A071102	Failure to Interrogate	06/06/2024	06/06/2024
									A071204	No Pacing	06/06/2024	06/06/2024
							F1901	Additional Surgery	A071102	Failure to Interrogate	06/06/2024	06/06/2024
									A071204	No Pacing	06/06/2024	06/06/2024
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A071102	Failure to Interrogate	06/06/2024	06/06/2024
									A071204	No Pacing	06/06/2024	06/06/2024
							F1901	Additional Surgery	A071102	Failure to Interrogate	06/06/2024	06/06/2024
									A071204	No Pacing	06/06/2024	06/06/2024
					CT2985	Active	E2403	No Clinical Signs, Symptoms or Conditions	A071102	Failure to Interrogate	06/06/2024	06/06/2024
									A071204	No Pacing	06/06/2024	06/06/2024
							F1901	Additional Surgery	A071102	Failure to Interrogate	06/06/2024	06/06/2024
									A071204	No Pacing	06/06/2024	06/06/2024
					CT3001	Implantable metal	E2403	No Clinical Signs, Symptoms or Conditions	A071102	Failure to Interrogate	06/06/2024	06/06/2024
									A071204	No Pacing	06/06/2024	06/06/2024
							F1901	Additional Surgery	A071102	Failure to Interrogate	06/06/2024	06/06/2024
									A071204	No Pacing	06/06/2024	06/06/2024
					CT3002	Implantable synthetic polymer	E2403	No Clinical Signs, Symptoms or Conditions	A071102	Failure to Interrogate	06/06/2024	06/06/2024
									A071204	No Pacing	06/06/2024	06/06/2024
							F1901	Additional Surgery	A071102	Failure to Interrogate	06/06/2024	06/06/2024
									A071204	No Pacing	06/06/2024	06/06/2024
					CT3006	Implantable inorganic material	E2403	No Clinical Signs, Symptoms or Conditions	A071102	Failure to Interrogate	06/06/2024	06/06/2024
									A071204	No Pacing	06/06/2024	06/06/2024
							F1901	Additional Surgery	A071102	Failure to Interrogate	06/06/2024	06/06/2024

									A071204	No Pacing	06/06/2024	06/06/2024
					CT314	Electrically powered	E2403	No Clinical Signs, Symptoms or Conditions	A071102	Failure to Interrogate	06/06/2024	06/06/2024
									A071204	No Pacing	06/06/2024	06/06/2024
							F1901	Additional Surgery	A071102	Failure to Interrogate	06/06/2024	06/06/2024
									A071204	No Pacing	06/06/2024	06/06/2024
					CT321	Long-term surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A071102	Failure to Interrogate	06/06/2024	06/06/2024
									A071204	No Pacing	06/06/2024	06/06/2024
							F1901	Additional Surgery	A071102	Failure to Interrogate	06/06/2024	06/06/2024
									A071204	No Pacing	06/06/2024	06/06/2024
					CT449	Generators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A071102	Failure to Interrogate	06/06/2024	06/06/2024
									A071204	No Pacing	06/06/2024	06/06/2024
							F1901	Additional Surgery	A071102	Failure to Interrogate	06/06/2024	06/06/2024
									A071204	No Pacing	06/06/2024	06/06/2024
					CT981	Single use	E2403	No Clinical Signs, Symptoms or Conditions	A071102	Failure to Interrogate	06/06/2024	06/06/2024
									A071204	No Pacing	06/06/2024	06/06/2024
							F1901	Additional Surgery	A071102	Failure to Interrogate	06/06/2024	06/06/2024
									A071204	No Pacing	06/06/2024	06/06/2024
					CT983	Surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A071102	Failure to Interrogate	06/06/2024	06/06/2024
									A071204	No Pacing	06/06/2024	06/06/2024
							F1901	Additional Surgery	A071102	Failure to Interrogate	06/06/2024	06/06/2024
									A071204	No Pacing	06/06/2024	06/06/2024
15791	2024	All other reportable incidents	35852	Single-chamber implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A160602	Premature End-of-Life Indicator		
							F1905	Device Revision or Replacement	A160602	Premature End-of-Life Indicator		
15792	2024	Unanticipated serious deterioration in state of health	47265	Dual-chamber implantable pacemaker, rate-responsive	CT1685	Cardiac pacemakers	E011903	Syncope/Fainting	A0712	Pacing Problem	10/08/2024	10/08/2024
							E060101	Asystole	A0712	Pacing Problem	10/08/2024	10/08/2024
							E0602	Cardiac Arrest	A0712	Pacing Problem	10/08/2024	10/08/2024
							F08	Hospitalization or Prolonged Hospitalization	A0712	Pacing Problem	10/08/2024	10/08/2024
							F1203	Life Threatening Illness or Injury	A0712	Pacing Problem	10/08/2024	10/08/2024
							F1905	Device Revision or Replacement	A0712	Pacing Problem	10/08/2024	10/08/2024
					CT257	Implantable cardiac pacemakers	E011903	Syncope/Fainting	A0712	Pacing Problem	10/08/2024	10/08/2024
							E060101	Asystole	A0712	Pacing Problem	10/08/2024	10/08/2024
							E0602	Cardiac Arrest	A0712	Pacing Problem	10/08/2024	10/08/2024
							F08	Hospitalization or Prolonged Hospitalization	A0712	Pacing Problem	10/08/2024	10/08/2024
							F1203	Life Threatening Illness or Injury	A0712	Pacing Problem	10/08/2024	10/08/2024
							F1905	Device Revision or Replacement	A0712	Pacing Problem	10/08/2024	10/08/2024
					CT2866	Implantable cardiac electrical stimulators	E011903	Syncope/Fainting	A0712	Pacing Problem	10/08/2024	10/08/2024
							E060101	Asystole	A0712	Pacing Problem	10/08/2024	10/08/2024
							E0602	Cardiac Arrest	A0712	Pacing Problem	10/08/2024	10/08/2024
							F08	Hospitalization or Prolonged Hospitalization	A0712	Pacing Problem	10/08/2024	10/08/2024
							F1203	Life Threatening Illness or Injury	A0712	Pacing Problem	10/08/2024	10/08/2024
							F1905	Device Revision or Replacement	A0712	Pacing Problem	10/08/2024	10/08/2024
15793	2024	All other reportable incidents	47265	Dual-chamber implantable pacemaker, rate-responsive	CT141	Electrodes/Leads and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070910	Undersensing		
									A090801	Signal Artifact/Noise		
							F26	No Health Consequences or Impact	A070910	Undersensing		
									A090801	Signal Artifact/Noise		
					CT145	Cardiovascular devices	E2403	No Clinical Signs, Symptoms or Conditions	A070910	Undersensing		
									A090801	Signal Artifact/Noise		
							F26	No Health Consequences or Impact	A070910	Undersensing		
									A090801	Signal Artifact/Noise		
					CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070910	Undersensing		
									A090801	Signal Artifact/Noise		
							F26	No Health Consequences or Impact	A070910	Undersensing		
									A090801	Signal Artifact/Noise		
					CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A070910	Undersensing		
									A090801	Signal Artifact/Noise		
							F26	No Health Consequences or Impact	A070910	Undersensing		
									A090801	Signal Artifact/Noise		
					CT233	Surgical	E2403	No Clinical Signs, Symptoms or Conditions	A070910	Undersensing		
									A090801	Signal Artifact/Noise		
							F26	No Health Consequences or Impact	A070910	Undersensing		
									A090801	Signal Artifact/Noise		
					CT2406	Implantable	E2403	No Clinical Signs, Symptoms or Conditions	A070910	Undersensing		
									A090801	Signal Artifact/Noise		
							F26	No Health Consequences or Impact	A070910	Undersensing		
									A090801	Signal Artifact/Noise		
					CT2407	Active implantable	E2403	No Clinical Signs, Symptoms or Conditions	A070910	Undersensing		
									A090801	Signal Artifact/Noise		
							F26	No Health Consequences or Impact	A070910	Undersensing		
									A090801	Signal Artifact/Noise		
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A070910	Undersensing		
									A090801	Signal Artifact/Noise		
							F26	No Health Consequences or Impact	A070910	Undersensing		
									A090801	Signal Artifact/Noise		
					CT2727	Active implant devices/ancillaries	E2403	No Clinical Signs, Symptoms or Conditions	A070910	Undersensing		
									A090801	Signal Artifact/Noise		
							F26	No Health Consequences or Impact	A070910	Undersensing		
									A090801	Signal Artifact/Noise		
					CT2750	Implantable cardiac electrical stimulators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070910	Undersensing		
									A090801	Signal Artifact/Noise		
							F26	No Health Consequences or Impact	A070910	Undersensing		
									A090801	Signal Artifact/Noise		
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A070910	Undersensing		
									A090801	Signal Artifact/Noise		
							F26	No Health Consequences or Impact	A070910	Undersensing		
									A090801	Signal Artifact/Noise		
					CT2985	Active	E2403	No Clinical Signs, Symptoms or Conditions	A070910	Undersensing		
									A090801	Signal Artifact/Noise		
							F26	No Health Consequences or Impact	A070910	Undersensing		

									A090801	Signal Artifact/Noise		
					CT3001	Implantable metal	E2403	No Clinical Signs, Symptoms or Conditions	A070910	Undersensing		
									A090801	Signal Artifact/Noise		
							F26	No Health Consequences or Impact	A070910	Undersensing		
									A090801	Signal Artifact/Noise		
					CT3002	Implantable synthetic polymer	E2403	No Clinical Signs, Symptoms or Conditions	A070910	Undersensing		
									A090801	Signal Artifact/Noise		
							F26	No Health Consequences or Impact	A070910	Undersensing		
									A090801	Signal Artifact/Noise		
					CT3006	Implantable inorganic material	E2403	No Clinical Signs, Symptoms or Conditions	A070910	Undersensing		
									A090801	Signal Artifact/Noise		
							F26	No Health Consequences or Impact	A070910	Undersensing		
									A090801	Signal Artifact/Noise		
					CT314	Electrically powered	E2403	No Clinical Signs, Symptoms or Conditions	A070910	Undersensing		
									A090801	Signal Artifact/Noise		
							F26	No Health Consequences or Impact	A070910	Undersensing		
									A090801	Signal Artifact/Noise		
					CT321	Long-term surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A070910	Undersensing		
									A090801	Signal Artifact/Noise		
							F26	No Health Consequences or Impact	A070910	Undersensing		
									A090801	Signal Artifact/Noise		
					CT449	Generators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070910	Undersensing		
									A090801	Signal Artifact/Noise		
							F26	No Health Consequences or Impact	A070910	Undersensing		
									A090801	Signal Artifact/Noise		
					CT981	Single use	E2403	No Clinical Signs, Symptoms or Conditions	A070910	Undersensing		
									A090801	Signal Artifact/Noise		
							F26	No Health Consequences or Impact	A070910	Undersensing		
									A090801	Signal Artifact/Noise		
					CT983	Surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A070910	Undersensing		
									A090801	Signal Artifact/Noise		
							F26	No Health Consequences or Impact	A070910	Undersensing		
									A090801	Signal Artifact/Noise		
15794	2024	All other reportable incidents	35852	Single-chamber implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A1605	Reset Problem		
									A160503	Inappropriate or Unexpected Reset		
							F26	No Health Consequences or Impact	A1605	Reset Problem		
									A160503	Inappropriate or Unexpected Reset		
15795	2024	All other reportable incidents	47270	Cardiac resynchronization therapy implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A070912	Incorrect Interpretation of Signal		
									A071302	Inappropriate/Inadequate Shock/Stimulation		
							F12	Serious Injury/ Illness/ Impairment	A070912	Incorrect Interpretation of Signal		
									A071302	Inappropriate/Inadequate Shock/Stimulation		
15796	2024	All other reportable incidents	17882	Physiologic-monitoring defibrillation system	CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070603	Failure to Charge		
							F27	Problem Identified During Non-Clinical Procedure	A070603	Failure to Charge		
15797	2024	All other reportable incidents	37265	Dual-chamber implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E060109	Tachycardia	A0709	Device Sensing Problem		
									A0712	Pacing Problem		
									A071301	Failure to Deliver Shock/Stimulation		
							F1001	Absence of Treatment	A0709	Device Sensing Problem		
									A0712	Pacing Problem		
									A071301	Failure to Deliver Shock/Stimulation		
							F1203	Life Threatening Illness or Injury	A0709	Device Sensing Problem		
									A0712	Pacing Problem		
									A071301	Failure to Deliver Shock/Stimulation		
							F2307	Reprogramming of Device	A0709	Device Sensing Problem		
									A0712	Pacing Problem		
									A071301	Failure to Deliver Shock/Stimulation		
15798	2024	All other reportable incidents	47265	Dual-chamber implantable pacemaker, rate-responsive	CT1685	Cardiac pacemakers	E1906	Unspecified Infection	A1304	Telemetry Discrepancy	23/02/2024	23/02/2024
							F08	Hospitalization or Prolonged Hospitalization	A1304	Telemetry Discrepancy	23/02/2024	23/02/2024
							F1905	Device Revision or Replacement	A1304	Telemetry Discrepancy	23/02/2024	23/02/2024
					CT257	Implantable cardiac pacemakers	E1906	Unspecified Infection	A1304	Telemetry Discrepancy	23/02/2024	23/02/2024
							F08	Hospitalization or Prolonged Hospitalization	A1304	Telemetry Discrepancy	23/02/2024	23/02/2024
							F1905	Device Revision or Replacement	A1304	Telemetry Discrepancy	23/02/2024	23/02/2024
					CT2866	Implantable cardiac electrical stimulators	E1906	Unspecified Infection	A1304	Telemetry Discrepancy	23/02/2024	23/02/2024
							F08	Hospitalization or Prolonged Hospitalization	A1304	Telemetry Discrepancy	23/02/2024	23/02/2024
							F1905	Device Revision or Replacement	A1304	Telemetry Discrepancy	23/02/2024	23/02/2024
15799	2024	Death	37265	Dual-chamber implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A0203	Defective Device	27/08/2024	27/08/2024
							F02	Death	A0203	Defective Device	27/08/2024	27/08/2024
15800	2024	All other reportable incidents	17882	Physiologic-monitoring defibrillation system	CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070903	Failure to Analyze Signal		
							F26	No Health Consequences or Impact	A070903	Failure to Analyze Signal		
15801	2024	All other reportable incidents	47205	Cardiac pulse generator programmer	CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A0902	Display or Visual Feedback Problem		
							F27	Problem Identified During Non-Clinical Procedure	A0902	Display or Visual Feedback Problem		
15802	2024	All other reportable incidents	47205	Cardiac pulse generator programmer	CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A0508	Noise, Audible		
									A0902	Display or Visual Feedback Problem		
							F27	Problem Identified During Non-Clinical Procedure	A0508	Noise, Audible		
									A0902	Display or Visual Feedback Problem		
15803	2024	All other reportable incidents	47265	Dual-chamber implantable pacemaker, rate-responsive	CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A05	Mechanical Problem	07/06/2024	07/06/2024
							F26	No Health Consequences or Impact	A05	Mechanical Problem	07/06/2024	07/06/2024
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A05	Mechanical Problem	07/06/2024	07/06/2024
							F26	No Health Consequences or Impact	A05	Mechanical Problem	07/06/2024	07/06/2024
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A05	Mechanical Problem	07/06/2024	07/06/2024
							F26	No Health Consequences or Impact	A05	Mechanical Problem	07/06/2024	07/06/2024
15804	2024	All other reportable incidents	36106	Implantable pacemaker analyser, intraoperative	CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A071102	Failure to Interrogate		
									A13	Communication or Transmission Problem		
							F26	No Health Consequences or Impact	A071102	Failure to Interrogate		
									A13	Communication or Transmission Problem		
15805	2024	All other reportable incidents	36106	Implantable pacemaker analyser, intraoperative	CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A071102	Failure to Interrogate		
									A13	Communication or Transmission Problem		
							F26	No Health Consequences or Impact	A071102	Failure to Interrogate		
									A13	Communication or Transmission Problem		

15806	2024		17882	Physiologic-monitoring defibrillation system	CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A090103	No Audible Prompt/Feedback		
									A2302	Use of Incorrect Control/Treatment Settings		
							F27	Problem Identified During Non-Clinical Procedure	A090103	No Audible Prompt/Feedback		
									A2302	Use of Incorrect Control/Treatment Settings		
15807	2024		47265	Dual-chamber implantable pacemaker, rate-responsive	CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	16/08/2024	16/08/2024
							F1901	Additional Surgery	A24	No Apparent Device Problem	16/08/2024	16/08/2024
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	16/08/2024	16/08/2024
							F1901	Additional Surgery	A24	No Apparent Device Problem	16/08/2024	16/08/2024
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	16/08/2024	16/08/2024
							F1901	Additional Surgery	A24	No Apparent Device Problem	16/08/2024	16/08/2024
15808	2024	All other reportable incidents	17882	Physiologic-monitoring defibrillation system	CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A0719	Unexpected Shutdown		
							F27	Problem Identified During Non-Clinical Procedure	A0719	Unexpected Shutdown		
15809	2024	All other reportable incidents	17882	Physiologic-monitoring defibrillation system	CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070803	Failure to Power Up		
							F27	Problem Identified During Non-Clinical Procedure	A070803	Failure to Power Up		
15810	2024	All other reportable incidents	47265	Dual-chamber implantable pacemaker, rate-responsive	CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	20/05/2024	20/05/2024
							F1901	Additional Surgery	A070504	Premature Discharge of Battery	20/05/2024	20/05/2024
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	20/05/2024	20/05/2024
							F1901	Additional Surgery	A070504	Premature Discharge of Battery	20/05/2024	20/05/2024
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	20/05/2024	20/05/2024
							F1901	Additional Surgery	A070504	Premature Discharge of Battery	20/05/2024	20/05/2024
15811	2024	All other reportable incidents	37265	Dual-chamber implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A13	Communication or Transmission Problem		
							F12	Serious Injury/ Illness/ Impairment	A13	Communication or Transmission Problem		
							F2307	Reprogramming of Device	A13	Communication or Transmission Problem		
15812	2024	All other reportable incidents	47265	Dual-chamber implantable pacemaker, rate-responsive	CT1685	Cardiac pacemakers	E2401	Insufficient Information	A072201	High Impedance	20/08/2024	
							F1903	Device Explantation	A072201	High Impedance	20/08/2024	
					CT257	Implantable cardiac pacemakers	E2401	Insufficient Information	A072201	High Impedance	20/08/2024	
							F1903	Device Explantation	A072201	High Impedance	20/08/2024	
					CT2866	Implantable cardiac electrical stimulators	E2401	Insufficient Information	A072201	High Impedance	20/08/2024	
							F1903	Device Explantation	A072201	High Impedance	20/08/2024	
15813	2024	All other reportable incidents	61499	Vascular catheter introduction kit, steerable, single-use	CT1684	Cardiac pacemakers and associated devices	E2401	Insufficient Information	A040607	Material Too Rigid or Stiff		
							F08	Hospitalization or Prolonged Hospitalization	A040607	Material Too Rigid or Stiff		
							F14	Prolonged Episode of Care	A040607	Material Too Rigid or Stiff		
							F23	Unexpected Medical Intervention	A040607	Material Too Rigid or Stiff		
15814	2024	All other reportable incidents	37265	Dual-chamber implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A072201	High Impedance		
							F26	No Health Consequences or Impact	A072201	High Impedance		
15815	2024	All other reportable incidents	47263	Cardiac resynchronization therapy implantable pacemaker	CT1685	Cardiac pacemakers	E011903	Syncope/Fainting	A070909	Oversensing		
									A071204	No Pacing		
							F23	Unexpected Medical Intervention	A070909	Oversensing		
									A071204	No Pacing		
					CT257	Implantable cardiac pacemakers	E011903	Syncope/Fainting	A070909	Oversensing		
									A071204	No Pacing		
							F23	Unexpected Medical Intervention	A070909	Oversensing		
									A071204	No Pacing		
					CT2866	Implantable cardiac electrical stimulators	E011903	Syncope/Fainting	A070909	Oversensing		
									A071204	No Pacing		
							F23	Unexpected Medical Intervention	A070909	Oversensing		
									A071204	No Pacing		
15816	2024	All other reportable incidents			CT1684	Cardiac pacemakers and associated devices	E060101	Asystole	A070909	Oversensing		
									A0712	Pacing Problem		
									A090801	Signal Artifact/Noise		
							F1001	Absence of Treatment	A070909	Oversensing		
									A0712	Pacing Problem		
									A090801	Signal Artifact/Noise		
15817	2024	All other reportable incidents	47265	Dual-chamber implantable pacemaker, rate-responsive	CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A0712	Pacing Problem		
							F26	No Health Consequences or Impact	A0712	Pacing Problem		
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A0712	Pacing Problem		
							F26	No Health Consequences or Impact	A0712	Pacing Problem		
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A0712	Pacing Problem		
							F26	No Health Consequences or Impact	A0712	Pacing Problem		
15818	2024	All other reportable incidents	47263	Cardiac resynchronization therapy implantable pacemaker	CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem	21/08/2024	21/08/2024
									A070504	Premature Discharge of Battery	21/08/2024	21/08/2024
							F08	Hospitalization or Prolonged Hospitalization	A0705	Battery Problem	21/08/2024	21/08/2024
									A070504	Premature Discharge of Battery	21/08/2024	21/08/2024
							F1905	Device Revision or Replacement	A0705	Battery Problem	21/08/2024	21/08/2024
									A070504	Premature Discharge of Battery	21/08/2024	21/08/2024
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem	21/08/2024	21/08/2024
									A070504	Premature Discharge of Battery	21/08/2024	21/08/2024
							F08	Hospitalization or Prolonged Hospitalization	A0705	Battery Problem	21/08/2024	21/08/2024
									A070504	Premature Discharge of Battery	21/08/2024	21/08/2024
							F1905	Device Revision or Replacement	A0705	Battery Problem	21/08/2024	21/08/2024
									A070504	Premature Discharge of Battery	21/08/2024	21/08/2024
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem	21/08/2024	21/08/2024
									A070504	Premature Discharge of Battery	21/08/2024	21/08/2024
							F08	Hospitalization or Prolonged Hospitalization	A0705	Battery Problem	21/08/2024	21/08/2024
									A070504	Premature Discharge of Battery	21/08/2024	21/08/2024
							F1905	Device Revision or Replacement	A0705	Battery Problem	21/08/2024	21/08/2024
									A070504	Premature Discharge of Battery	21/08/2024	21/08/2024
15819	2024	All other reportable incidents	47265	Dual-chamber implantable pacemaker, rate-responsive	CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
15820	2024	All other reportable incidents	47270	Cardiac resynchronization therapy implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A070909	Oversensing		
									A0908	Incorrect, Inadequate or Imprecise Result or Readings		
							F26	No Health Consequences or Impact	A070909	Oversensing		

									A0908	Incorrect, Inadequate or Imprecise Result or Readings		
15821	2024		35224	External pacemaker, epicardial pacing	CT1685	Cardiac pacemakers	F26	No Health Consequences or Impact	A0202	Defective Component		
									A070503	Failure to Run on Battery		
15822	2024	All other reportable incidents	47205	Cardiac pulse generator programmer	CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A0902	Display or Visual Feedback Problem		
							F26	No Health Consequences or Impact	A0902	Display or Visual Feedback Problem		
15823	2024	All other reportable incidents	37265	Dual-chamber implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	21/10/2024	21/10/2024
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	21/10/2024	21/10/2024
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	21/10/2024	21/10/2024
15824	2024	All other reportable incidents	17882	Physiologic-monitoring defibrillation system	CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A0719	Unexpected Shutdown		
							F27	Problem Identified During Non-Clinical Procedure	A0719	Unexpected Shutdown		
15825	2024	All other reportable incidents	16038	Implantable pulse generator mesh bag, non-bioabsorbable	CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070909	Oversensing	23/05/2024	23/05/2024
							F08	Hospitalization or Prolonged Hospitalization	A070909	Oversensing	23/05/2024	23/05/2024
							F1905	Device Revision or Replacement	A070909	Oversensing	23/05/2024	23/05/2024
15826	2024	All other reportable incidents	37265	Dual-chamber implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem	18/03/2024	18/03/2024
									A070504	Premature Discharge of Battery	18/03/2024	18/03/2024
									A0908	Incorrect, Inadequate or Imprecise Result or Readings	18/03/2024	18/03/2024
							F08	Hospitalization or Prolonged Hospitalization	A0705	Battery Problem	18/03/2024	18/03/2024
									A070504	Premature Discharge of Battery	18/03/2024	18/03/2024
									A0908	Incorrect, Inadequate or Imprecise Result or Readings	18/03/2024	18/03/2024
							F1905	Device Revision or Replacement	A0705	Battery Problem	18/03/2024	18/03/2024
									A070504	Premature Discharge of Battery	18/03/2024	18/03/2024
									A0908	Incorrect, Inadequate or Imprecise Result or Readings	18/03/2024	18/03/2024
15827	2024	All other reportable incidents	47265	Dual-chamber implantable pacemaker, rate-responsive	CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A090801	Signal Artifact/Noise	23/08/2024	23/08/2024
									A1205	Loose or Intermittent Connection	23/08/2024	23/08/2024
							F1901	Additional Surgery	A090801	Signal Artifact/Noise	23/08/2024	23/08/2024
									A1205	Loose or Intermittent Connection	23/08/2024	23/08/2024
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A090801	Signal Artifact/Noise	23/08/2024	23/08/2024
									A1205	Loose or Intermittent Connection	23/08/2024	23/08/2024
							F1901	Additional Surgery	A090801	Signal Artifact/Noise	23/08/2024	23/08/2024
									A1205	Loose or Intermittent Connection	23/08/2024	23/08/2024
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A090801	Signal Artifact/Noise	23/08/2024	23/08/2024
									A1205	Loose or Intermittent Connection	23/08/2024	23/08/2024
							F1901	Additional Surgery	A090801	Signal Artifact/Noise	23/08/2024	23/08/2024
									A1205	Loose or Intermittent Connection	23/08/2024	23/08/2024
15828	2024	All other reportable incidents	47270	Cardiac resynchronization therapy implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
							F24	Insufficient Health Impact Information	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
15829	2024	All other reportable incidents	47263	Cardiac resynchronization therapy implantable pacemaker	CT1685	Cardiac pacemakers	E011903	Syncope/Fainting	A070101	Failure to Capture	04/10/2023	04/10/2023
									A0908	Incorrect, Inadequate or Imprecise Result or Readings	04/10/2023	04/10/2023
							E060104	Bradycardia	A070101	Failure to Capture	04/10/2023	04/10/2023
									A0908	Incorrect, Inadequate or Imprecise Result or Readings	04/10/2023	04/10/2023
							F08	Hospitalization or Prolonged Hospitalization	A070101	Failure to Capture	04/10/2023	04/10/2023
									A0908	Incorrect, Inadequate or Imprecise Result or Readings	04/10/2023	04/10/2023
							F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A070101	Failure to Capture	04/10/2023	04/10/2023
									A0908	Incorrect, Inadequate or Imprecise Result or Readings	04/10/2023	04/10/2023
							F1905	Device Revision or Replacement	A070101	Failure to Capture	04/10/2023	04/10/2023
									A0908	Incorrect, Inadequate or Imprecise Result or Readings	04/10/2023	04/10/2023
					CT257	Implantable cardiac pacemakers	E011903	Syncope/Fainting	A070101	Failure to Capture	04/10/2023	04/10/2023
									A0908	Incorrect, Inadequate or Imprecise Result or Readings	04/10/2023	04/10/2023
							E060104	Bradycardia	A070101	Failure to Capture	04/10/2023	04/10/2023
									A0908	Incorrect, Inadequate or Imprecise Result or Readings	04/10/2023	04/10/2023
							F08	Hospitalization or Prolonged Hospitalization	A070101	Failure to Capture	04/10/2023	04/10/2023
									A0908	Incorrect, Inadequate or Imprecise Result or Readings	04/10/2023	04/10/2023
							F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A070101	Failure to Capture	04/10/2023	04/10/2023
									A0908	Incorrect, Inadequate or Imprecise Result or Readings	04/10/2023	04/10/2023
							F1905	Device Revision or Replacement	A070101	Failure to Capture	04/10/2023	04/10/2023
									A0908	Incorrect, Inadequate or Imprecise Result or Readings	04/10/2023	04/10/2023
					CT2866	Implantable cardiac electrical stimulators	E011903	Syncope/Fainting	A070101	Failure to Capture	04/10/2023	04/10/2023
									A0908	Incorrect, Inadequate or Imprecise Result or Readings	04/10/2023	04/10/2023
							E060104	Bradycardia	A070101	Failure to Capture	04/10/2023	04/10/2023
									A0908	Incorrect, Inadequate or Imprecise Result or Readings	04/10/2023	04/10/2023
							F08	Hospitalization or Prolonged Hospitalization	A070101	Failure to Capture	04/10/2023	04/10/2023
									A0908	Incorrect, Inadequate or Imprecise Result or Readings	04/10/2023	04/10/2023
							F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A070101	Failure to Capture	04/10/2023	04/10/2023
									A0908	Incorrect, Inadequate or Imprecise Result or Readings	04/10/2023	04/10/2023
							F1905	Device Revision or Replacement	A070101	Failure to Capture	04/10/2023	04/10/2023
									A0908	Incorrect, Inadequate or Imprecise Result or Readings	04/10/2023	04/10/2023
15830	2024	Unanticipated serious deterioration in state of health	47265	Dual-chamber implantable pacemaker, rate-responsive	CT1685	Cardiac pacemakers	E0112	Dizziness	A05	Mechanical Problem	04/09/2024	04/09/2024
									A0712	Pacing Problem	04/09/2024	04/09/2024
							E011902	Presyncope	A05	Mechanical Problem	04/09/2024	04/09/2024
									A0712	Pacing Problem	04/09/2024	04/09/2024
							E0618	Non Specific EKG/ECG Changes	A05	Mechanical Problem	04/09/2024	04/09/2024

									A0712	Pacing Problem	04/09/2024	04/09/2024
							F08	Hospitalization or Prolonged Hospitalization	A05	Mechanical Problem	04/09/2024	04/09/2024
									A0712	Pacing Problem	04/09/2024	04/09/2024
							F1203	Life Threatening Illness or Injury	A05	Mechanical Problem	04/09/2024	04/09/2024
									A0712	Pacing Problem	04/09/2024	04/09/2024
							F1905	Device Revision or Replacement	A05	Mechanical Problem	04/09/2024	04/09/2024
									A0712	Pacing Problem	04/09/2024	04/09/2024
					CT257	Implantable cardiac pacemakers	E0112	Dizziness	A05	Mechanical Problem	04/09/2024	04/09/2024
									A0712	Pacing Problem	04/09/2024	04/09/2024
							E011902	Presyncope	A05	Mechanical Problem	04/09/2024	04/09/2024
									A0712	Pacing Problem	04/09/2024	04/09/2024
							E0618	Non Specific EKG/ECG Changes	A05	Mechanical Problem	04/09/2024	04/09/2024
									A0712	Pacing Problem	04/09/2024	04/09/2024
							F08	Hospitalization or Prolonged Hospitalization	A05	Mechanical Problem	04/09/2024	04/09/2024
									A0712	Pacing Problem	04/09/2024	04/09/2024
							F1203	Life Threatening Illness or Injury	A05	Mechanical Problem	04/09/2024	04/09/2024
									A0712	Pacing Problem	04/09/2024	04/09/2024
							F1905	Device Revision or Replacement	A05	Mechanical Problem	04/09/2024	04/09/2024
									A0712	Pacing Problem	04/09/2024	04/09/2024
					CT2866	Implantable cardiac electrical stimulators	E0112	Dizziness	A05	Mechanical Problem	04/09/2024	04/09/2024
									A0712	Pacing Problem	04/09/2024	04/09/2024
							E011902	Presyncope	A05	Mechanical Problem	04/09/2024	04/09/2024
									A0712	Pacing Problem	04/09/2024	04/09/2024
							E0618	Non Specific EKG/ECG Changes	A05	Mechanical Problem	04/09/2024	04/09/2024
									A0712	Pacing Problem	04/09/2024	04/09/2024
							F08	Hospitalization or Prolonged Hospitalization	A05	Mechanical Problem	04/09/2024	04/09/2024
									A0712	Pacing Problem	04/09/2024	04/09/2024
							F1203	Life Threatening Illness or Injury	A05	Mechanical Problem	04/09/2024	04/09/2024
									A0712	Pacing Problem	04/09/2024	04/09/2024
							F1905	Device Revision or Replacement	A05	Mechanical Problem	04/09/2024	04/09/2024
									A0712	Pacing Problem	04/09/2024	04/09/2024
15831	2024	All other reportable incidents	47205	Cardiac pulse generator programmer	CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A0902	Display or Visual Feedback Problem		
							F27	Problem Identified During Non-Clinical Procedure	A0902	Display or Visual Feedback Problem		
15832	2024	All other reportable incidents	47270	Cardiac resynchronization therapy implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E060109	Tachycardia	A0709	Device Sensing Problem		
									A071301	Failure to Deliver Shock/Stimulation		
							F1001	Absence of Treatment	A0709	Device Sensing Problem		
									A071301	Failure to Deliver Shock/Stimulation		
							F1203	Life Threatening Illness or Injury	A0709	Device Sensing Problem		
									A071301	Failure to Deliver Shock/Stimulation		
							F2307	Reprogramming of Device	A0709	Device Sensing Problem		
									A071301	Failure to Deliver Shock/Stimulation		
15833	2024	All other reportable incidents	60789	Intracardiac pacemaker	CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
									A090805	Incorrect Measurement		
									A1605	Reset Problem		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
									A090805	Incorrect Measurement		
									A1605	Reset Problem		
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
									A090805	Incorrect Measurement		
									A1605	Reset Problem		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
									A090805	Incorrect Measurement		
									A1605	Reset Problem		
					CT2818	Catheterization/Cannulation/Fluid path devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
									A090805	Incorrect Measurement		
									A1605	Reset Problem		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
									A090805	Incorrect Measurement		
									A1605	Reset Problem		
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
									A090805	Incorrect Measurement		
									A1605	Reset Problem		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
									A090805	Incorrect Measurement		
									A1605	Reset Problem		
15834	2024	All other reportable incidents	35852	Single-chamber implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A0709	Device Sensing Problem		
									A0712	Pacing Problem		
							F26	No Health Consequences or Impact	A0709	Device Sensing Problem		
									A0712	Pacing Problem		
15835	2024	All other reportable incidents	47270	Cardiac resynchronization therapy implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A070909	Oversensing		
									A072001	Electromagnetic Interference		
									A090801	Signal Artifact/Noise		
									A160105	False Alarm		
							F26	No Health Consequences or Impact	A070909	Oversensing		
									A072001	Electromagnetic Interference		
									A090801	Signal Artifact/Noise		
									A160105	False Alarm		
15836	2024	All other reportable incidents	47265	Dual-chamber implantable pacemaker, rate-responsive	CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem		
							F1901	Additional Surgery	A24	No Apparent Device Problem		
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem		
							F1901	Additional Surgery	A24	No Apparent Device Problem		
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem		
							F1901	Additional Surgery	A24	No Apparent Device Problem		
15837	2024	All other reportable incidents	47265	Dual-chamber implantable pacemaker, rate-responsive	CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	03/03/2024	03/03/2024
							F1901	Additional Surgery	A24	No Apparent Device Problem	03/03/2024	03/03/2024
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	03/03/2024	03/03/2024
							F1901	Additional Surgery	A24	No Apparent Device Problem	03/03/2024	03/03/2024
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	03/03/2024	03/03/2024

							F1901	Additional Surgery	A24	No Apparent Device Problem	03/03/2024	03/03/2024
15838	2024	All other reportable incidents	17882	Physiologic-monitoring defibrillation system	CT1684	Cardiac pacemakers and associated devices			A0501	Detachment of Device or Device Component		
15839	2024	All other reportable incidents	47205	Cardiac pulse generator programmer	CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A0902	Display or Visual Feedback Problem		
							F27	Problem Identified During Non-Clinical Procedure	A0902	Display or Visual Feedback Problem		
15840	2024	All other reportable incidents	37265	Dual-chamber implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A0401	Break		
									A150201	Positioning Failure		
									A23	Use of Device Problem		
							F05	Delay to Treatment/ Therapy	A0401	Break		
									A150201	Positioning Failure		
									A23	Use of Device Problem		
15841	2024	All other reportable incidents	37265	Dual-chamber implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A05	Mechanical Problem		
									A072201	High Impedance		
							F26	No Health Consequences or Impact	A05	Mechanical Problem		
									A072201	High Impedance		
15842	2024	All other reportable incidents	47265	Dual-chamber implantable pacemaker, rate-responsive	CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A071102	Failure to Interrogate	26/01/2024	26/01/2024
									A071204	No Pacing	26/01/2024	26/01/2024
							F1901	Additional Surgery	A071102	Failure to Interrogate	26/01/2024	26/01/2024
									A071204	No Pacing	26/01/2024	26/01/2024
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A071102	Failure to Interrogate	26/01/2024	26/01/2024
									A071204	No Pacing	26/01/2024	26/01/2024
							F1901	Additional Surgery	A071102	Failure to Interrogate	26/01/2024	26/01/2024
									A071204	No Pacing	26/01/2024	26/01/2024
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A071102	Failure to Interrogate	26/01/2024	26/01/2024
									A071204	No Pacing	26/01/2024	26/01/2024
							F1901	Additional Surgery	A071102	Failure to Interrogate	26/01/2024	26/01/2024
									A071204	No Pacing	26/01/2024	26/01/2024
15843	2024		37265	Dual-chamber implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E0601	Arrhythmia	A25	No Apparent Adverse Event	26/06/2024	26/06/2024
							F02	Death	A25	No Apparent Adverse Event	26/06/2024	26/06/2024
15844	2024	All other reportable incidents	47263	Cardiac resynchronization therapy implantable pacemaker	CT1685	Cardiac pacemakers	E2103	Device Overstimulation of Tissue	A0908	Incorrect, Inadequate or Imprecise Result or Readings	23/04/2024	23/04/2024
							F08	Hospitalization or Prolonged Hospitalization	A0908	Incorrect, Inadequate or Imprecise Result or Readings	23/04/2024	23/04/2024
							F15	Recognized Device or Procedural Complication	A0908	Incorrect, Inadequate or Imprecise Result or Readings	23/04/2024	23/04/2024
							F1905	Device Revision or Replacement	A0908	Incorrect, Inadequate or Imprecise Result or Readings	23/04/2024	23/04/2024
					CT257	Implantable cardiac pacemakers	E2103	Device Overstimulation of Tissue	A0908	Incorrect, Inadequate or Imprecise Result or Readings	23/04/2024	23/04/2024
							F08	Hospitalization or Prolonged Hospitalization	A0908	Incorrect, Inadequate or Imprecise Result or Readings	23/04/2024	23/04/2024
							F15	Recognized Device or Procedural Complication	A0908	Incorrect, Inadequate or Imprecise Result or Readings	23/04/2024	23/04/2024
							F1905	Device Revision or Replacement	A0908	Incorrect, Inadequate or Imprecise Result or Readings	23/04/2024	23/04/2024
					CT2866	Implantable cardiac electrical stimulators	E2103	Device Overstimulation of Tissue	A0908	Incorrect, Inadequate or Imprecise Result or Readings	23/04/2024	23/04/2024
							F08	Hospitalization or Prolonged Hospitalization	A0908	Incorrect, Inadequate or Imprecise Result or Readings	23/04/2024	23/04/2024
							F15	Recognized Device or Procedural Complication	A0908	Incorrect, Inadequate or Imprecise Result or Readings	23/04/2024	23/04/2024
							F1905	Device Revision or Replacement	A0908	Incorrect, Inadequate or Imprecise Result or Readings	23/04/2024	23/04/2024
15845	2024	All other reportable incidents	47265	Dual-chamber implantable pacemaker, rate-responsive	CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	02/05/2024	02/05/2024
									A0711	Interrogation Problem	02/05/2024	02/05/2024
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	02/05/2024	02/05/2024
									A0711	Interrogation Problem	02/05/2024	02/05/2024
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	02/05/2024	02/05/2024
									A0711	Interrogation Problem	02/05/2024	02/05/2024
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	02/05/2024	02/05/2024
									A0711	Interrogation Problem	02/05/2024	02/05/2024
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	02/05/2024	02/05/2024
									A0711	Interrogation Problem	02/05/2024	02/05/2024
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	02/05/2024	02/05/2024
									A0711	Interrogation Problem	02/05/2024	02/05/2024
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	02/05/2024	02/05/2024
									A0711	Interrogation Problem	02/05/2024	02/05/2024
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	02/05/2024	02/05/2024
									A0711	Interrogation Problem	02/05/2024	02/05/2024
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	02/05/2024	02/05/2024
									A0711	Interrogation Problem	02/05/2024	02/05/2024
15846	2024	All other reportable incidents	47265	Dual-chamber implantable pacemaker, rate-responsive	CT141	Electrodes/Leads and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem	07/02/2024	07/02/2024
									A070504	Premature Discharge of Battery	07/02/2024	07/02/2024
							F08	Hospitalization or Prolonged Hospitalization	A0705	Battery Problem	07/02/2024	07/02/2024
									A070504	Premature Discharge of Battery	07/02/2024	07/02/2024
							F1905	Device Revision or Replacement	A0705	Battery Problem	07/02/2024	07/02/2024
									A070504	Premature Discharge of Battery	07/02/2024	07/02/2024
					CT145	Cardiovascular devices	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem	07/02/2024	07/02/2024
									A070504	Premature Discharge of Battery	07/02/2024	07/02/2024
							F08	Hospitalization or Prolonged Hospitalization	A0705	Battery Problem	07/02/2024	07/02/2024
									A070504	Premature Discharge of Battery	07/02/2024	07/02/2024
							F1905	Device Revision or Replacement	A0705	Battery Problem	07/02/2024	07/02/2024
									A070504	Premature Discharge of Battery	07/02/2024	07/02/2024
					CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem	07/02/2024	07/02/2024
									A070504	Premature Discharge of Battery	07/02/2024	07/02/2024
							F08	Hospitalization or Prolonged Hospitalization	A0705	Battery Problem	07/02/2024	07/02/2024
									A070504	Premature Discharge of Battery	07/02/2024	07/02/2024
							F1905	Device Revision or Replacement	A0705	Battery Problem	07/02/2024	07/02/2024
									A070504	Premature Discharge of Battery	07/02/2024	07/02/2024
					CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem	07/02/2024	07/02/2024
									A070504	Premature Discharge of Battery	07/02/2024	07/02/2024
							F08	Hospitalization or Prolonged Hospitalization	A0705	Battery Problem	07/02/2024	07/02/2024

						A070504	Premature Discharge of Battery	07/02/2024	07/02/2024
					F1905	Device Revision or Replacement	A0705	Battery Problem	07/02/2024	07/02/2024
							A070504	Premature Discharge of Battery	07/02/2024	07/02/2024
				CT233	Surgical	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem	07/02/2024	07/02/2024
								A070504	Premature Discharge of Battery	07/02/2024	07/02/2024
						F08	Hospitalization or Prolonged Hospitalization	A0705	Battery Problem	07/02/2024	07/02/2024
								A070504	Premature Discharge of Battery	07/02/2024	07/02/2024
						F1905	Device Revision or Replacement	A0705	Battery Problem	07/02/2024	07/02/2024
								A070504	Premature Discharge of Battery	07/02/2024	07/02/2024
				CT2406	Implantable	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem	07/02/2024	07/02/2024
								A070504	Premature Discharge of Battery	07/02/2024	07/02/2024
						F08	Hospitalization or Prolonged Hospitalization	A0705	Battery Problem	07/02/2024	07/02/2024
								A070504	Premature Discharge of Battery	07/02/2024	07/02/2024
						F1905	Device Revision or Replacement	A0705	Battery Problem	07/02/2024	07/02/2024
								A070504	Premature Discharge of Battery	07/02/2024	07/02/2024
				CT2407	Active implantable	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem	07/02/2024	07/02/2024
								A070504	Premature Discharge of Battery	07/02/2024	07/02/2024
						F08	Hospitalization or Prolonged Hospitalization	A0705	Battery Problem	07/02/2024	07/02/2024
								A070504	Premature Discharge of Battery	07/02/2024	07/02/2024
						F1905	Device Revision or Replacement	A0705	Battery Problem	07/02/2024	07/02/2024
								A070504	Premature Discharge of Battery	07/02/2024	07/02/2024
				CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem	07/02/2024	07/02/2024
								A070504	Premature Discharge of Battery	07/02/2024	07/02/2024
						F08	Hospitalization or Prolonged Hospitalization	A0705	Battery Problem	07/02/2024	07/02/2024
								A070504	Premature Discharge of Battery	07/02/2024	07/02/2024
						F1905	Device Revision or Replacement	A0705	Battery Problem	07/02/2024	07/02/2024
								A070504	Premature Discharge of Battery	07/02/2024	07/02/2024
				CT2727	Active implant devices/ancillaries	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem	07/02/2024	07/02/2024
								A070504	Premature Discharge of Battery	07/02/2024	07/02/2024
						F08	Hospitalization or Prolonged Hospitalization	A0705	Battery Problem	07/02/2024	07/02/2024
								A070504	Premature Discharge of Battery	07/02/2024	07/02/2024
						F1905	Device Revision or Replacement	A0705	Battery Problem	07/02/2024	07/02/2024
								A070504	Premature Discharge of Battery	07/02/2024	07/02/2024
				CT2750	Implantable cardiac electrical stimulators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem	07/02/2024	07/02/2024
								A070504	Premature Discharge of Battery	07/02/2024	07/02/2024
						F08	Hospitalization or Prolonged Hospitalization	A0705	Battery Problem	07/02/2024	07/02/2024
								A070504	Premature Discharge of Battery	07/02/2024	07/02/2024
						F1905	Device Revision or Replacement	A0705	Battery Problem	07/02/2024	07/02/2024
								A070504	Premature Discharge of Battery	07/02/2024	07/02/2024
				CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem	07/02/2024	07/02/2024
								A070504	Premature Discharge of Battery	07/02/2024	07/02/2024
						F08	Hospitalization or Prolonged Hospitalization	A0705	Battery Problem	07/02/2024	07/02/2024
								A070504	Premature Discharge of Battery	07/02/2024	07/02/2024
						F1905	Device Revision or Replacement	A0705	Battery Problem	07/02/2024	07/02/2024
								A070504	Premature Discharge of Battery	07/02/2024	07/02/2024
				CT2985	Active	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem	07/02/2024	07/02/2024
								A070504	Premature Discharge of Battery	07/02/2024	07/02/2024
						F08	Hospitalization or Prolonged Hospitalization	A0705	Battery Problem	07/02/2024	07/02/2024
								A070504	Premature Discharge of Battery	07/02/2024	07/02/2024
						F1905	Device Revision or Replacement	A0705	Battery Problem	07/02/2024	07/02/2024
								A070504	Premature Discharge of Battery	07/02/2024	07/02/2024
				CT3001	Implantable metal	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem	07/02/2024	07/02/2024
								A070504	Premature Discharge of Battery	07/02/2024	07/02/2024
						F08	Hospitalization or Prolonged Hospitalization	A0705	Battery Problem	07/02/2024	07/02/2024
								A070504	Premature Discharge of Battery	07/02/2024	07/02/2024
						F1905	Device Revision or Replacement	A0705	Battery Problem	07/02/2024	07/02/2024
								A070504	Premature Discharge of Battery	07/02/2024	07/02/2024
				CT3002	Implantable synthetic polymer	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem	07/02/2024	07/02/2024
								A070504	Premature Discharge of Battery	07/02/2024	07/02/2024
						F08	Hospitalization or Prolonged Hospitalization	A0705	Battery Problem	07/02/2024	07/02/2024
								A070504	Premature Discharge of Battery	07/02/2024	07/02/2024
						F1905	Device Revision or Replacement	A0705	Battery Problem	07/02/2024	07/02/2024
								A070504	Premature Discharge of Battery	07/02/2024	07/02/2024
				CT3006	Implantable inorganic material	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem	07/02/2024	07/02/2024
								A070504	Premature Discharge of Battery	07/02/2024	07/02/2024
						F08	Hospitalization or Prolonged Hospitalization	A0705	Battery Problem	07/02/2024	07/02/2024
								A070504	Premature Discharge of Battery	07/02/2024	07/02/2024
						F1905	Device Revision or Replacement	A0705	Battery Problem	07/02/2024	07/02/2024
								A070504	Premature Discharge of Battery	07/02/2024	07/02/2024
				CT314	Electrically powered	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem	07/02/2024	07/02/2024
								A070504	Premature Discharge of Battery	07/02/2024	07/02/2024
						F08	Hospitalization or Prolonged Hospitalization	A0705	Battery Problem	07/02/2024	07/02/2024
								A070504	Premature Discharge of Battery	07/02/2024	07/02/2024
						F1905	Device Revision or Replacement	A0705	Battery Problem	07/02/2024	07/02/2024
								A070504	Premature Discharge of Battery	07/02/2024	07/02/2024
				CT321	Long-term surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem	07/02/2024	07/02/2024
								A070504	Premature Discharge of Battery	07/02/2024	07/02/2024
						F08	Hospitalization or Prolonged Hospitalization	A0705	Battery Problem	07/02/2024	07/02/2024
								A070504	Premature Discharge of Battery	07/02/2024	07/02/2024
						F1905	Device Revision or Replacement	A0705	Battery Problem	07/02/2024	07/02/2024
								A070504	Premature Discharge of Battery	07/02/2024	07/02/2024
				CT449	Generators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem	07/02/2024	07/02/2024
								A070504	Premature Discharge of Battery	07/02/2024	07/02/2024
						F08	Hospitalization or Prolonged Hospitalization	A0705	Battery Problem	07/02/2024	07/02/2024
								A070504	Premature Discharge of Battery	07/02/2024	07/02/2024
						F1905	Device Revision or Replacement	A0705	Battery Problem	07/02/2024	07/02/2024
								A070504	Premature Discharge of Battery	07/02/2024	07/02/2024
				CT981	Single use	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem	07/02/2024	07/02/2024

									A070504	Premature Discharge of Battery	07/02/2024	07/02/2024
							F08	Hospitalization or Prolonged Hospitalization	A0705	Battery Problem	07/02/2024	07/02/2024
									A070504	Premature Discharge of Battery	07/02/2024	07/02/2024
							F1905	Device Revision or Replacement	A0705	Battery Problem	07/02/2024	07/02/2024
									A070504	Premature Discharge of Battery	07/02/2024	07/02/2024
					CT983	Surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem	07/02/2024	07/02/2024
									A070504	Premature Discharge of Battery	07/02/2024	07/02/2024
							F08	Hospitalization or Prolonged Hospitalization	A0705	Battery Problem	07/02/2024	07/02/2024
									A070504	Premature Discharge of Battery	07/02/2024	07/02/2024
							F1905	Device Revision or Replacement	A0705	Battery Problem	07/02/2024	07/02/2024
									A070504	Premature Discharge of Battery	07/02/2024	07/02/2024
15847	2024	All other reportable incidents	47270	Cardiac resynchronization therapy implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	19/11/2024	19/11/2024
									A0908	Incorrect, Inadequate or Imprecise Result or Readings	19/11/2024	19/11/2024
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	19/11/2024	19/11/2024
									A0908	Incorrect, Inadequate or Imprecise Result or Readings	19/11/2024	19/11/2024
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	19/11/2024	19/11/2024
									A0908	Incorrect, Inadequate or Imprecise Result or Readings	19/11/2024	19/11/2024
15848	2024	All other reportable incidents	47265	Dual-chamber implantable pacemaker, rate-responsive	CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
15849	2024	All other reportable incidents	47267	Single-chamber implantable pacemaker, rate-responsive	CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A0712	Pacing Problem	22/03/2024	22/03/2024
							F26	No Health Consequences or Impact	A0712	Pacing Problem	22/03/2024	22/03/2024
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A0712	Pacing Problem	22/03/2024	22/03/2024
							F26	No Health Consequences or Impact	A0712	Pacing Problem	22/03/2024	22/03/2024
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A0712	Pacing Problem	22/03/2024	22/03/2024
							F26	No Health Consequences or Impact	A0712	Pacing Problem	22/03/2024	22/03/2024
15850	2024	All other reportable incidents	47265	Dual-chamber implantable pacemaker, rate-responsive	CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A0908	Incorrect, Inadequate or Imprecise Result or Readings		
							F26	No Health Consequences or Impact	A0908	Incorrect, Inadequate or Imprecise Result or Readings		
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A0908	Incorrect, Inadequate or Imprecise Result or Readings		
							F26	No Health Consequences or Impact	A0908	Incorrect, Inadequate or Imprecise Result or Readings		
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A0908	Incorrect, Inadequate or Imprecise Result or Readings		
							F26	No Health Consequences or Impact	A0908	Incorrect, Inadequate or Imprecise Result or Readings		
15851	2024	All other reportable incidents	47265	Dual-chamber implantable pacemaker, rate-responsive	CT141	Electrodes/Leads and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A071102	Failure to Interrogate	27/06/2024	27/06/2024
									A071204	No Pacing	27/06/2024	27/06/2024
							F1901	Additional Surgery	A071102	Failure to Interrogate	27/06/2024	27/06/2024
									A071204	No Pacing	27/06/2024	27/06/2024
					CT145	Cardiovascular devices	E2403	No Clinical Signs, Symptoms or Conditions	A071102	Failure to Interrogate	27/06/2024	27/06/2024
									A071204	No Pacing	27/06/2024	27/06/2024
							F1901	Additional Surgery	A071102	Failure to Interrogate	27/06/2024	27/06/2024
									A071204	No Pacing	27/06/2024	27/06/2024
					CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A071102	Failure to Interrogate	27/06/2024	27/06/2024
									A071204	No Pacing	27/06/2024	27/06/2024
							F1901	Additional Surgery	A071102	Failure to Interrogate	27/06/2024	27/06/2024
									A071204	No Pacing	27/06/2024	27/06/2024
					CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A071102	Failure to Interrogate	27/06/2024	27/06/2024
									A071204	No Pacing	27/06/2024	27/06/2024
							F1901	Additional Surgery	A071102	Failure to Interrogate	27/06/2024	27/06/2024
									A071204	No Pacing	27/06/2024	27/06/2024
					CT233	Surgical	E2403	No Clinical Signs, Symptoms or Conditions	A071102	Failure to Interrogate	27/06/2024	27/06/2024
									A071204	No Pacing	27/06/2024	27/06/2024
							F1901	Additional Surgery	A071102	Failure to Interrogate	27/06/2024	27/06/2024
									A071204	No Pacing	27/06/2024	27/06/2024
					CT2406	Implantable	E2403	No Clinical Signs, Symptoms or Conditions	A071102	Failure to Interrogate	27/06/2024	27/06/2024
									A071204	No Pacing	27/06/2024	27/06/2024
							F1901	Additional Surgery	A071102	Failure to Interrogate	27/06/2024	27/06/2024
									A071204	No Pacing	27/06/2024	27/06/2024
					CT2407	Active implantable	E2403	No Clinical Signs, Symptoms or Conditions	A071102	Failure to Interrogate	27/06/2024	27/06/2024
									A071204	No Pacing	27/06/2024	27/06/2024
							F1901	Additional Surgery	A071102	Failure to Interrogate	27/06/2024	27/06/2024
									A071204	No Pacing	27/06/2024	27/06/2024
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A071102	Failure to Interrogate	27/06/2024	27/06/2024
									A071204	No Pacing	27/06/2024	27/06/2024
							F1901	Additional Surgery	A071102	Failure to Interrogate	27/06/2024	27/06/2024
									A071204	No Pacing	27/06/2024	27/06/2024
					CT2727	Active implant devices/ancillaries	E2403	No Clinical Signs, Symptoms or Conditions	A071102	Failure to Interrogate	27/06/2024	27/06/2024
									A071204	No Pacing	27/06/2024	27/06/2024
							F1901	Additional Surgery	A071102	Failure to Interrogate	27/06/2024	27/06/2024
									A071204	No Pacing	27/06/2024	27/06/2024
					CT2750	Implantable cardiac electrical stimulators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A071102	Failure to Interrogate	27/06/2024	27/06/2024
									A071204	No Pacing	27/06/2024	27/06/2024
							F1901	Additional Surgery	A071102	Failure to Interrogate	27/06/2024	27/06/2024
									A071204	No Pacing	27/06/2024	27/06/2024
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A071102	Failure to Interrogate	27/06/2024	27/06/2024
									A071204	No Pacing	27/06/2024	27/06/2024
							F1901	Additional Surgery	A071102	Failure to Interrogate	27/06/2024	27/06/2024
									A071204	No Pacing	27/06/2024	27/06/2024
					CT2985	Active	E2403	No Clinical Signs, Symptoms or Conditions	A071102	Failure to Interrogate	27/06/2024	27/06/2024
									A071204	No Pacing	27/06/2024	27/06/2024

							F1901	Additional Surgery	A071102	Failure to Interrogate	27/06/2024	27/06/2024
									A071204	No Pacing	27/06/2024	27/06/2024
					CT3001	Implantable metal	E2403	No Clinical Signs, Symptoms or Conditions	A071102	Failure to Interrogate	27/06/2024	27/06/2024
									A071204	No Pacing	27/06/2024	27/06/2024
							F1901	Additional Surgery	A071102	Failure to Interrogate	27/06/2024	27/06/2024
									A071204	No Pacing	27/06/2024	27/06/2024
					CT3002	Implantable synthetic polymer	E2403	No Clinical Signs, Symptoms or Conditions	A071102	Failure to Interrogate	27/06/2024	27/06/2024
									A071204	No Pacing	27/06/2024	27/06/2024
							F1901	Additional Surgery	A071102	Failure to Interrogate	27/06/2024	27/06/2024
									A071204	No Pacing	27/06/2024	27/06/2024
					CT3006	Implantable inorganic material	E2403	No Clinical Signs, Symptoms or Conditions	A071102	Failure to Interrogate	27/06/2024	27/06/2024
									A071204	No Pacing	27/06/2024	27/06/2024
							F1901	Additional Surgery	A071102	Failure to Interrogate	27/06/2024	27/06/2024
									A071204	No Pacing	27/06/2024	27/06/2024
					CT314	Electrically powered	E2403	No Clinical Signs, Symptoms or Conditions	A071102	Failure to Interrogate	27/06/2024	27/06/2024
									A071204	No Pacing	27/06/2024	27/06/2024
							F1901	Additional Surgery	A071102	Failure to Interrogate	27/06/2024	27/06/2024
									A071204	No Pacing	27/06/2024	27/06/2024
					CT321	Long-term surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A071102	Failure to Interrogate	27/06/2024	27/06/2024
									A071204	No Pacing	27/06/2024	27/06/2024
							F1901	Additional Surgery	A071102	Failure to Interrogate	27/06/2024	27/06/2024
									A071204	No Pacing	27/06/2024	27/06/2024
					CT449	Generators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A071102	Failure to Interrogate	27/06/2024	27/06/2024
									A071204	No Pacing	27/06/2024	27/06/2024
							F1901	Additional Surgery	A071102	Failure to Interrogate	27/06/2024	27/06/2024
									A071204	No Pacing	27/06/2024	27/06/2024
					CT981	Single use	E2403	No Clinical Signs, Symptoms or Conditions	A071102	Failure to Interrogate	27/06/2024	27/06/2024
									A071204	No Pacing	27/06/2024	27/06/2024
							F1901	Additional Surgery	A071102	Failure to Interrogate	27/06/2024	27/06/2024
									A071204	No Pacing	27/06/2024	27/06/2024
					CT983	Surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A071102	Failure to Interrogate	27/06/2024	27/06/2024
									A071204	No Pacing	27/06/2024	27/06/2024
							F1901	Additional Surgery	A071102	Failure to Interrogate	27/06/2024	27/06/2024
									A071204	No Pacing	27/06/2024	27/06/2024
15852	2024	All other reportable incidents	47265	Dual-chamber implantable pacemaker, rate-responsive	CT141	Electrodes/Leads and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070909	Oversensing	06/06/2024	06/06/2024
									A071204	No Pacing	06/06/2024	06/06/2024
							F1901	Additional Surgery	A070909	Oversensing	06/06/2024	06/06/2024
									A071204	No Pacing	06/06/2024	06/06/2024
					CT145	Cardiovascular devices	E2403	No Clinical Signs, Symptoms or Conditions	A070909	Oversensing	06/06/2024	06/06/2024
									A071204	No Pacing	06/06/2024	06/06/2024
							F1901	Additional Surgery	A070909	Oversensing	06/06/2024	06/06/2024
									A071204	No Pacing	06/06/2024	06/06/2024
					CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070909	Oversensing	06/06/2024	06/06/2024
									A071204	No Pacing	06/06/2024	06/06/2024
							F1901	Additional Surgery	A070909	Oversensing	06/06/2024	06/06/2024
									A071204	No Pacing	06/06/2024	06/06/2024
					CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A070909	Oversensing	06/06/2024	06/06/2024
									A071204	No Pacing	06/06/2024	06/06/2024
							F1901	Additional Surgery	A070909	Oversensing	06/06/2024	06/06/2024
									A071204	No Pacing	06/06/2024	06/06/2024
					CT233	Surgical	E2403	No Clinical Signs, Symptoms or Conditions	A070909	Oversensing	06/06/2024	06/06/2024
									A071204	No Pacing	06/06/2024	06/06/2024
							F1901	Additional Surgery	A070909	Oversensing	06/06/2024	06/06/2024
									A071204	No Pacing	06/06/2024	06/06/2024
					CT2406	Implantable	E2403	No Clinical Signs, Symptoms or Conditions	A070909	Oversensing	06/06/2024	06/06/2024
									A071204	No Pacing	06/06/2024	06/06/2024
							F1901	Additional Surgery	A070909	Oversensing	06/06/2024	06/06/2024
									A071204	No Pacing	06/06/2024	06/06/2024
					CT2407	Active implantable	E2403	No Clinical Signs, Symptoms or Conditions	A070909	Oversensing	06/06/2024	06/06/2024
									A071204	No Pacing	06/06/2024	06/06/2024
							F1901	Additional Surgery	A070909	Oversensing	06/06/2024	06/06/2024
									A071204	No Pacing	06/06/2024	06/06/2024
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A070909	Oversensing	06/06/2024	06/06/2024
									A071204	No Pacing	06/06/2024	06/06/2024
							F1901	Additional Surgery	A070909	Oversensing	06/06/2024	06/06/2024
									A071204	No Pacing	06/06/2024	06/06/2024
					CT2727	Active implant devices/ancillaries	E2403	No Clinical Signs, Symptoms or Conditions	A070909	Oversensing	06/06/2024	06/06/2024
									A071204	No Pacing	06/06/2024	06/06/2024
							F1901	Additional Surgery	A070909	Oversensing	06/06/2024	06/06/2024
									A071204	No Pacing	06/06/2024	06/06/2024
					CT2750	Implantable cardiac electrical stimulators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070909	Oversensing	06/06/2024	06/06/2024
									A071204	No Pacing	06/06/2024	06/06/2024
							F1901	Additional Surgery	A070909	Oversensing	06/06/2024	06/06/2024
									A071204	No Pacing	06/06/2024	06/06/2024
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A070909	Oversensing	06/06/2024	06/06/2024
									A071204	No Pacing	06/06/2024	06/06/2024
							F1901	Additional Surgery	A070909	Oversensing	06/06/2024	06/06/2024
									A071204	No Pacing	06/06/2024	06/06/2024
					CT2985	Active	E2403	No Clinical Signs, Symptoms or Conditions	A070909	Oversensing	06/06/2024	06/06/2024
									A071204	No Pacing	06/06/2024	06/06/2024
							F1901	Additional Surgery	A070909	Oversensing	06/06/2024	06/06/2024
									A071204	No Pacing	06/06/2024	06/06/2024
					CT3001	Implantable metal	E2403	No Clinical Signs, Symptoms or Conditions	A070909	Oversensing	06/06/2024	06/06/2024
									A071204	No Pacing	06/06/2024	06/06/2024
							F1901	Additional Surgery	A070909	Oversensing	06/06/2024	06/06/2024
									A071204	No Pacing	06/06/2024	06/06/2024
					CT3002	Implantable synthetic polymer	E2403	No Clinical Signs, Symptoms or Conditions	A070909	Oversensing	06/06/2024	06/06/2024
									A071204	No Pacing	06/06/2024	06/06/2024

						F1901	Additional Surgery	A070909	Oversensing	06/06/2024	06/06/2024
								A071204	No Pacing	06/06/2024	06/06/2024
					CT3006	Implantable inorganic material	E2403	No Clinical Signs, Symptoms or Conditions	A070909	Oversensing	06/06/2024	06/06/2024
									A071204	No Pacing	06/06/2024	06/06/2024
							F1901	Additional Surgery	A070909	Oversensing	06/06/2024	06/06/2024
									A071204	No Pacing	06/06/2024	06/06/2024
					CT314	Electrically powered	E2403	No Clinical Signs, Symptoms or Conditions	A070909	Oversensing	06/06/2024	06/06/2024
									A071204	No Pacing	06/06/2024	06/06/2024
							F1901	Additional Surgery	A070909	Oversensing	06/06/2024	06/06/2024
									A071204	No Pacing	06/06/2024	06/06/2024
					CT321	Long-term surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A070909	Oversensing	06/06/2024	06/06/2024
									A071204	No Pacing	06/06/2024	06/06/2024
							F1901	Additional Surgery	A070909	Oversensing	06/06/2024	06/06/2024
									A071204	No Pacing	06/06/2024	06/06/2024
					CT449	Generators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070909	Oversensing	06/06/2024	06/06/2024
									A071204	No Pacing	06/06/2024	06/06/2024
							F1901	Additional Surgery	A070909	Oversensing	06/06/2024	06/06/2024
									A071204	No Pacing	06/06/2024	06/06/2024
					CT981	Single use	E2403	No Clinical Signs, Symptoms or Conditions	A070909	Oversensing	06/06/2024	06/06/2024
									A071204	No Pacing	06/06/2024	06/06/2024
							F1901	Additional Surgery	A070909	Oversensing	06/06/2024	06/06/2024
									A071204	No Pacing	06/06/2024	06/06/2024
					CT983	Surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A070909	Oversensing	06/06/2024	06/06/2024
									A071204	No Pacing	06/06/2024	06/06/2024
							F1901	Additional Surgery	A070909	Oversensing	06/06/2024	06/06/2024
									A071204	No Pacing	06/06/2024	06/06/2024
15853	2024	Death	37265	Dual-chamber implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E0601	Arrhythmia	A070912	Incorrect Interpretation of Signal		
									A071201	Failure to Convert Rhythm		
									A071302	Inappropriate/Inadequate Shock/Stimulation		
							F02	Death	A070912	Incorrect Interpretation of Signal		
									A071201	Failure to Convert Rhythm		
									A071302	Inappropriate/Inadequate Shock/Stimulation		
15854	2024	All other reportable incidents	47265	Dual-chamber implantable pacemaker, rate-responsive	CT141	Electrodes/Leads and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem	22/08/2024	22/08/2024
									A070504	Premature Discharge of Battery	22/08/2024	22/08/2024
							F08	Hospitalization or Prolonged Hospitalization	A0705	Battery Problem	22/08/2024	22/08/2024
									A070504	Premature Discharge of Battery	22/08/2024	22/08/2024
							F1905	Device Revision or Replacement	A0705	Battery Problem	22/08/2024	22/08/2024
									A070504	Premature Discharge of Battery	22/08/2024	22/08/2024
					CT145	Cardiovascular devices	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem	22/08/2024	22/08/2024
									A070504	Premature Discharge of Battery	22/08/2024	22/08/2024
							F08	Hospitalization or Prolonged Hospitalization	A0705	Battery Problem	22/08/2024	22/08/2024
									A070504	Premature Discharge of Battery	22/08/2024	22/08/2024
							F1905	Device Revision or Replacement	A0705	Battery Problem	22/08/2024	22/08/2024
									A070504	Premature Discharge of Battery	22/08/2024	22/08/2024
					CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem	22/08/2024	22/08/2024
									A070504	Premature Discharge of Battery	22/08/2024	22/08/2024
							F08	Hospitalization or Prolonged Hospitalization	A0705	Battery Problem	22/08/2024	22/08/2024
									A070504	Premature Discharge of Battery	22/08/2024	22/08/2024
							F1905	Device Revision or Replacement	A0705	Battery Problem	22/08/2024	22/08/2024
									A070504	Premature Discharge of Battery	22/08/2024	22/08/2024
					CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem	22/08/2024	22/08/2024
									A070504	Premature Discharge of Battery	22/08/2024	22/08/2024
							F08	Hospitalization or Prolonged Hospitalization	A0705	Battery Problem	22/08/2024	22/08/2024
									A070504	Premature Discharge of Battery	22/08/2024	22/08/2024
							F1905	Device Revision or Replacement	A0705	Battery Problem	22/08/2024	22/08/2024
									A070504	Premature Discharge of Battery	22/08/2024	22/08/2024
					CT233	Surgical	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem	22/08/2024	22/08/2024
									A070504	Premature Discharge of Battery	22/08/2024	22/08/2024
							F08	Hospitalization or Prolonged Hospitalization	A0705	Battery Problem	22/08/2024	22/08/2024
									A070504	Premature Discharge of Battery	22/08/2024	22/08/2024
							F1905	Device Revision or Replacement	A0705	Battery Problem	22/08/2024	22/08/2024
									A070504	Premature Discharge of Battery	22/08/2024	22/08/2024
					CT2406	Implantable	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem	22/08/2024	22/08/2024
									A070504	Premature Discharge of Battery	22/08/2024	22/08/2024
							F08	Hospitalization or Prolonged Hospitalization	A0705	Battery Problem	22/08/2024	22/08/2024
									A070504	Premature Discharge of Battery	22/08/2024	22/08/2024
							F1905	Device Revision or Replacement	A0705	Battery Problem	22/08/2024	22/08/2024
									A070504	Premature Discharge of Battery	22/08/2024	22/08/2024
					CT2407	Active implantable	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem	22/08/2024	22/08/2024
									A070504	Premature Discharge of Battery	22/08/2024	22/08/2024
							F08	Hospitalization or Prolonged Hospitalization	A0705	Battery Problem	22/08/2024	22/08/2024
									A070504	Premature Discharge of Battery	22/08/2024	22/08/2024
							F1905	Device Revision or Replacement	A0705	Battery Problem	22/08/2024	22/08/2024
									A070504	Premature Discharge of Battery	22/08/2024	22/08/2024
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem	22/08/2024	22/08/2024
									A070504	Premature Discharge of Battery	22/08/2024	22/08/2024
							F08	Hospitalization or Prolonged Hospitalization	A0705	Battery Problem	22/08/2024	22/08/2024
									A070504	Premature Discharge of Battery	22/08/2024	22/08/2024
							F1905	Device Revision or Replacement	A0705	Battery Problem	22/08/2024	22/08/2024
									A070504	Premature Discharge of Battery	22/08/2024	22/08/2024
					CT2727	Active implant devices/ancillaries	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem	22/08/2024	22/08/2024
									A070504	Premature Discharge of Battery	22/08/2024	22/08/2024
							F08	Hospitalization or Prolonged Hospitalization	A0705	Battery Problem	22/08/2024	22/08/2024
									A070504	Premature Discharge of Battery	22/08/2024	22/08/2024
							F1905	Device Revision or Replacement	A0705	Battery Problem	22/08/2024	22/08/2024
									A070504	Premature Discharge of Battery	22/08/2024	22/08/2024
					CT2750	Implantable cardiac electrical stimulators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem	22/08/2024	22/08/2024
									A070504	Premature Discharge of Battery	22/08/2024	22/08/2024

							F08	Hospitalization or Prolonged Hospitalization	A0705	Battery Problem	22/08/2024	22/08/2024
									A070504	Premature Discharge of Battery	22/08/2024	22/08/2024
							F1905	Device Revision or Replacement	A0705	Battery Problem	22/08/2024	22/08/2024
									A070504	Premature Discharge of Battery	22/08/2024	22/08/2024
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem	22/08/2024	22/08/2024
									A070504	Premature Discharge of Battery	22/08/2024	22/08/2024
							F08	Hospitalization or Prolonged Hospitalization	A0705	Battery Problem	22/08/2024	22/08/2024
									A070504	Premature Discharge of Battery	22/08/2024	22/08/2024
							F1905	Device Revision or Replacement	A0705	Battery Problem	22/08/2024	22/08/2024
									A070504	Premature Discharge of Battery	22/08/2024	22/08/2024
					CT2985	Active	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem	22/08/2024	22/08/2024
									A070504	Premature Discharge of Battery	22/08/2024	22/08/2024
							F08	Hospitalization or Prolonged Hospitalization	A0705	Battery Problem	22/08/2024	22/08/2024
									A070504	Premature Discharge of Battery	22/08/2024	22/08/2024
							F1905	Device Revision or Replacement	A0705	Battery Problem	22/08/2024	22/08/2024
									A070504	Premature Discharge of Battery	22/08/2024	22/08/2024
					CT3001	Implantable metal	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem	22/08/2024	22/08/2024
									A070504	Premature Discharge of Battery	22/08/2024	22/08/2024
							F08	Hospitalization or Prolonged Hospitalization	A0705	Battery Problem	22/08/2024	22/08/2024
									A070504	Premature Discharge of Battery	22/08/2024	22/08/2024
							F1905	Device Revision or Replacement	A0705	Battery Problem	22/08/2024	22/08/2024
									A070504	Premature Discharge of Battery	22/08/2024	22/08/2024
					CT3002	Implantable synthetic polymer	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem	22/08/2024	22/08/2024
									A070504	Premature Discharge of Battery	22/08/2024	22/08/2024
							F08	Hospitalization or Prolonged Hospitalization	A0705	Battery Problem	22/08/2024	22/08/2024
									A070504	Premature Discharge of Battery	22/08/2024	22/08/2024
							F1905	Device Revision or Replacement	A0705	Battery Problem	22/08/2024	22/08/2024
									A070504	Premature Discharge of Battery	22/08/2024	22/08/2024
					CT3006	Implantable inorganic material	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem	22/08/2024	22/08/2024
									A070504	Premature Discharge of Battery	22/08/2024	22/08/2024
							F08	Hospitalization or Prolonged Hospitalization	A0705	Battery Problem	22/08/2024	22/08/2024
									A070504	Premature Discharge of Battery	22/08/2024	22/08/2024
							F1905	Device Revision or Replacement	A0705	Battery Problem	22/08/2024	22/08/2024
									A070504	Premature Discharge of Battery	22/08/2024	22/08/2024
					CT314	Electrically powered	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem	22/08/2024	22/08/2024
									A070504	Premature Discharge of Battery	22/08/2024	22/08/2024
							F08	Hospitalization or Prolonged Hospitalization	A0705	Battery Problem	22/08/2024	22/08/2024
									A070504	Premature Discharge of Battery	22/08/2024	22/08/2024
							F1905	Device Revision or Replacement	A0705	Battery Problem	22/08/2024	22/08/2024
									A070504	Premature Discharge of Battery	22/08/2024	22/08/2024
					CT321	Long-term surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem	22/08/2024	22/08/2024
									A070504	Premature Discharge of Battery	22/08/2024	22/08/2024
							F08	Hospitalization or Prolonged Hospitalization	A0705	Battery Problem	22/08/2024	22/08/2024
									A070504	Premature Discharge of Battery	22/08/2024	22/08/2024
							F1905	Device Revision or Replacement	A0705	Battery Problem	22/08/2024	22/08/2024
									A070504	Premature Discharge of Battery	22/08/2024	22/08/2024
					CT449	Generators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem	22/08/2024	22/08/2024
									A070504	Premature Discharge of Battery	22/08/2024	22/08/2024
							F08	Hospitalization or Prolonged Hospitalization	A0705	Battery Problem	22/08/2024	22/08/2024
									A070504	Premature Discharge of Battery	22/08/2024	22/08/2024
							F1905	Device Revision or Replacement	A0705	Battery Problem	22/08/2024	22/08/2024
									A070504	Premature Discharge of Battery	22/08/2024	22/08/2024
					CT981	Single use	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem	22/08/2024	22/08/2024
									A070504	Premature Discharge of Battery	22/08/2024	22/08/2024
							F08	Hospitalization or Prolonged Hospitalization	A0705	Battery Problem	22/08/2024	22/08/2024
									A070504	Premature Discharge of Battery	22/08/2024	22/08/2024
							F1905	Device Revision or Replacement	A0705	Battery Problem	22/08/2024	22/08/2024
									A070504	Premature Discharge of Battery	22/08/2024	22/08/2024
					CT983	Surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem	22/08/2024	22/08/2024
									A070504	Premature Discharge of Battery	22/08/2024	22/08/2024
							F08	Hospitalization or Prolonged Hospitalization	A0705	Battery Problem	22/08/2024	22/08/2024
									A070504	Premature Discharge of Battery	22/08/2024	22/08/2024
							F1905	Device Revision or Replacement	A0705	Battery Problem	22/08/2024	22/08/2024
									A070504	Premature Discharge of Battery	22/08/2024	22/08/2024
15855	2024	All other reportable incidents	47263	Cardiac resynchronization therapy implantable pacemaker	CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	25/06/2024	25/06/2024
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	25/06/2024	25/06/2024
							F12	Serious Injury/ Illness/ Impairment	A070504	Premature Discharge of Battery	25/06/2024	25/06/2024
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	25/06/2024	25/06/2024
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	25/06/2024	25/06/2024
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	25/06/2024	25/06/2024
							F12	Serious Injury/ Illness/ Impairment	A070504	Premature Discharge of Battery	25/06/2024	25/06/2024
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	25/06/2024	25/06/2024
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	25/06/2024	25/06/2024
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	25/06/2024	25/06/2024
							F12	Serious Injury/ Illness/ Impairment	A070504	Premature Discharge of Battery	25/06/2024	25/06/2024
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	25/06/2024	25/06/2024
15856	2024	All other reportable incidents	47265	Dual-chamber implantable pacemaker, rate-responsive	CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A010402	Migration		
									A1304	Telemetry Discrepancy		
							F2203	Imaging Required	A010402	Migration		
									A1304	Telemetry Discrepancy		
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A010402	Migration		
									A1304	Telemetry Discrepancy		
							F2203	Imaging Required	A010402	Migration		
									A1304	Telemetry Discrepancy		
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A010402	Migration		
									A1304	Telemetry Discrepancy		
							F2203	Imaging Required	A010402	Migration		
									A1304	Telemetry Discrepancy		

15857	2024	All other reportable incidents	47265	Dual-chamber implantable pacemaker, rate-responsive	CT1685	Cardiac pacemakers	E0623	Unspecified Heart Problem	A070909	Oversensing		
							E0717	Dyspnea	A070909	Oversensing		
							F23	Unexpected Medical Intervention	A070909	Oversensing		
					CT257	Implantable cardiac pacemakers	E0623	Unspecified Heart Problem	A070909	Oversensing		
							E0717	Dyspnea	A070909	Oversensing		
							F23	Unexpected Medical Intervention	A070909	Oversensing		
					CT2866	Implantable cardiac electrical stimulators	E0623	Unspecified Heart Problem	A070909	Oversensing		
							E0717	Dyspnea	A070909	Oversensing		
							F23	Unexpected Medical Intervention	A070909	Oversensing		
15858	2024	All other reportable incidents	47265	Dual-chamber implantable pacemaker, rate-responsive	CT141	Electrodes/Leads and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	02/04/2024	02/04/2024
							F1901	Additional Surgery	A24	No Apparent Device Problem	02/04/2024	02/04/2024
					CT145	Cardiovascular devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	02/04/2024	02/04/2024
							F1901	Additional Surgery	A24	No Apparent Device Problem	02/04/2024	02/04/2024
					CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	02/04/2024	02/04/2024
							F1901	Additional Surgery	A24	No Apparent Device Problem	02/04/2024	02/04/2024
					CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	02/04/2024	02/04/2024
							F1901	Additional Surgery	A24	No Apparent Device Problem	02/04/2024	02/04/2024
					CT233	Surgical	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	02/04/2024	02/04/2024
							F1901	Additional Surgery	A24	No Apparent Device Problem	02/04/2024	02/04/2024
					CT2406	Implantable	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	02/04/2024	02/04/2024
							F1901	Additional Surgery	A24	No Apparent Device Problem	02/04/2024	02/04/2024
					CT2407	Active implantable	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	02/04/2024	02/04/2024
							F1901	Additional Surgery	A24	No Apparent Device Problem	02/04/2024	02/04/2024
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	02/04/2024	02/04/2024
							F1901	Additional Surgery	A24	No Apparent Device Problem	02/04/2024	02/04/2024
					CT2727	Active implant devices/ancillaries	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	02/04/2024	02/04/2024
							F1901	Additional Surgery	A24	No Apparent Device Problem	02/04/2024	02/04/2024
					CT2750	Implantable cardiac electrical stimulators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	02/04/2024	02/04/2024
							F1901	Additional Surgery	A24	No Apparent Device Problem	02/04/2024	02/04/2024
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	02/04/2024	02/04/2024
							F1901	Additional Surgery	A24	No Apparent Device Problem	02/04/2024	02/04/2024
					CT2985	Active	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	02/04/2024	02/04/2024
							F1901	Additional Surgery	A24	No Apparent Device Problem	02/04/2024	02/04/2024
					CT3001	Implantable metal	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	02/04/2024	02/04/2024
							F1901	Additional Surgery	A24	No Apparent Device Problem	02/04/2024	02/04/2024
					CT3002	Implantable synthetic polymer	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	02/04/2024	02/04/2024
							F1901	Additional Surgery	A24	No Apparent Device Problem	02/04/2024	02/04/2024
					CT3006	Implantable inorganic material	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	02/04/2024	02/04/2024
							F1901	Additional Surgery	A24	No Apparent Device Problem	02/04/2024	02/04/2024
					CT314	Electrically powered	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	02/04/2024	02/04/2024
							F1901	Additional Surgery	A24	No Apparent Device Problem	02/04/2024	02/04/2024
					CT321	Long-term surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	02/04/2024	02/04/2024
							F1901	Additional Surgery	A24	No Apparent Device Problem	02/04/2024	02/04/2024
					CT449	Generators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	02/04/2024	02/04/2024
							F1901	Additional Surgery	A24	No Apparent Device Problem	02/04/2024	02/04/2024
					CT981	Single use	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	02/04/2024	02/04/2024
							F1901	Additional Surgery	A24	No Apparent Device Problem	02/04/2024	02/04/2024
					CT983	Surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	02/04/2024	02/04/2024
							F1901	Additional Surgery	A24	No Apparent Device Problem	02/04/2024	02/04/2024
15859	2024		47265	Dual-chamber implantable pacemaker, rate-responsive	CT141	Electrodes/Leads and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	29/05/2024	29/05/2024
							F1901	Additional Surgery	A24	No Apparent Device Problem	29/05/2024	29/05/2024
					CT145	Cardiovascular devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	29/05/2024	29/05/2024
							F1901	Additional Surgery	A24	No Apparent Device Problem	29/05/2024	29/05/2024
					CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	29/05/2024	29/05/2024
							F1901	Additional Surgery	A24	No Apparent Device Problem	29/05/2024	29/05/2024
					CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	29/05/2024	29/05/2024
							F1901	Additional Surgery	A24	No Apparent Device Problem	29/05/2024	29/05/2024
					CT233	Surgical	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	29/05/2024	29/05/2024
							F1901	Additional Surgery	A24	No Apparent Device Problem	29/05/2024	29/05/2024
					CT2406	Implantable	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	29/05/2024	29/05/2024
							F1901	Additional Surgery	A24	No Apparent Device Problem	29/05/2024	29/05/2024
					CT2407	Active implantable	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	29/05/2024	29/05/2024
							F1901	Additional Surgery	A24	No Apparent Device Problem	29/05/2024	29/05/2024
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	29/05/2024	29/05/2024
							F1901	Additional Surgery	A24	No Apparent Device Problem	29/05/2024	29/05/2024
					CT2727	Active implant devices/ancillaries	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	29/05/2024	29/05/2024
							F1901	Additional Surgery	A24	No Apparent Device Problem	29/05/2024	29/05/2024
					CT2750	Implantable cardiac electrical stimulators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	29/05/2024	29/05/2024
							F1901	Additional Surgery	A24	No Apparent Device Problem	29/05/2024	29/05/2024
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	29/05/2024	29/05/2024
							F1901	Additional Surgery	A24	No Apparent Device Problem	29/05/2024	29/05/2024
					CT2985	Active	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	29/05/2024	29/05/2024
							F1901	Additional Surgery	A24	No Apparent Device Problem	29/05/2024	29/05/2024
					CT3001	Implantable metal	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	29/05/2024	29/05/2024
							F1901	Additional Surgery	A24	No Apparent Device Problem	29/05/2024	29/05/2024
					CT3002	Implantable synthetic polymer	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	29/05/2024	29/05/2024
							F1901	Additional Surgery	A24	No Apparent Device Problem	29/05/2024	29/05/2024
					CT3006	Implantable inorganic material	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	29/05/2024	29/05/2024
							F1901	Additional Surgery	A24	No Apparent Device Problem	29/05/2024	29/05/2024
					CT314	Electrically powered	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	29/05/2024	29/05/2024
							F1901	Additional Surgery	A24	No Apparent Device Problem	29/05/2024	29/05/2024
					CT321	Long-term surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	29/05/2024	29/05/2024
							F1901	Additional Surgery	A24	No Apparent Device Problem	29/05/2024	29/05/2024
					CT449	Generators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	29/05/2024	29/05/2024
							F1901	Additional Surgery	A24	No Apparent Device Problem	29/05/2024	29/05/2024
					CT981	Single use	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	29/05/2024	29/05/2024
							F1901	Additional Surgery	A24	No Apparent Device Problem	29/05/2024	29/05/2024
					CT983	Surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	29/05/2024	29/05/2024

							F1901	Additional Surgery	A24	No Apparent Device Problem	29/05/2024	29/05/2024
15860	2024	All other reportable incidents	47267	Single-chamber implantable pacemaker, rate-responsive	CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A0712	Pacing Problem	22/03/2024	22/03/2024
							F26	No Health Consequences or Impact	A0712	Pacing Problem	22/03/2024	22/03/2024
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A0712	Pacing Problem	22/03/2024	22/03/2024
							F26	No Health Consequences or Impact	A0712	Pacing Problem	22/03/2024	22/03/2024
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A0712	Pacing Problem	22/03/2024	22/03/2024
							F26	No Health Consequences or Impact	A0712	Pacing Problem	22/03/2024	22/03/2024
15861	2024	All other reportable incidents	47265	Dual-chamber implantable pacemaker, rate-responsive	CT1685	Cardiac pacemakers	E011903	Syncope/Fainting	A0908	Incorrect, Inadequate or Imprecise Result or Readings	06/09/2024	06/09/2024
							E060101	Asystole	A0908	Incorrect, Inadequate or Imprecise Result or Readings	06/09/2024	06/09/2024
							F08	Hospitalization or Prolonged Hospitalization	A0908	Incorrect, Inadequate or Imprecise Result or Readings	06/09/2024	06/09/2024
							F1905	Device Revision or Replacement	A0908	Incorrect, Inadequate or Imprecise Result or Readings	06/09/2024	06/09/2024
							F2303	Medication Required	A0908	Incorrect, Inadequate or Imprecise Result or Readings	06/09/2024	06/09/2024
					CT257	Implantable cardiac pacemakers	E011903	Syncope/Fainting	A0908	Incorrect, Inadequate or Imprecise Result or Readings	06/09/2024	06/09/2024
							E060101	Asystole	A0908	Incorrect, Inadequate or Imprecise Result or Readings	06/09/2024	06/09/2024
							F08	Hospitalization or Prolonged Hospitalization	A0908	Incorrect, Inadequate or Imprecise Result or Readings	06/09/2024	06/09/2024
							F1905	Device Revision or Replacement	A0908	Incorrect, Inadequate or Imprecise Result or Readings	06/09/2024	06/09/2024
							F2303	Medication Required	A0908	Incorrect, Inadequate or Imprecise Result or Readings	06/09/2024	06/09/2024
					CT2866	Implantable cardiac electrical stimulators	E011903	Syncope/Fainting	A0908	Incorrect, Inadequate or Imprecise Result or Readings	06/09/2024	06/09/2024
							E060101	Asystole	A0908	Incorrect, Inadequate or Imprecise Result or Readings	06/09/2024	06/09/2024
							F08	Hospitalization or Prolonged Hospitalization	A0908	Incorrect, Inadequate or Imprecise Result or Readings	06/09/2024	06/09/2024
							F1905	Device Revision or Replacement	A0908	Incorrect, Inadequate or Imprecise Result or Readings	06/09/2024	06/09/2024
							F2303	Medication Required	A0908	Incorrect, Inadequate or Imprecise Result or Readings	06/09/2024	06/09/2024
15862	2024		35852	Single-chamber implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem		
									A160601	Premature Elective Replacement Indicator		
									A160602	Premature End-of-Life Indicator		
							F26	No Health Consequences or Impact	A0705	Battery Problem		
									A160601	Premature Elective Replacement Indicator		
									A160602	Premature End-of-Life Indicator		
15863	2024	All other reportable incidents	47263	Cardiac resynchronization therapy implantable pacemaker	CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A0908	Incorrect, Inadequate or Imprecise Result or Readings	22/12/2023	22/12/2023
							F08	Hospitalization or Prolonged Hospitalization	A0908	Incorrect, Inadequate or Imprecise Result or Readings	22/12/2023	22/12/2023
							F1905	Device Revision or Replacement	A0908	Incorrect, Inadequate or Imprecise Result or Readings	22/12/2023	22/12/2023
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A0908	Incorrect, Inadequate or Imprecise Result or Readings	22/12/2023	22/12/2023
							F08	Hospitalization or Prolonged Hospitalization	A0908	Incorrect, Inadequate or Imprecise Result or Readings	22/12/2023	22/12/2023
							F1905	Device Revision or Replacement	A0908	Incorrect, Inadequate or Imprecise Result or Readings	22/12/2023	22/12/2023
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A0908	Incorrect, Inadequate or Imprecise Result or Readings	22/12/2023	22/12/2023
							F08	Hospitalization or Prolonged Hospitalization	A0908	Incorrect, Inadequate or Imprecise Result or Readings	22/12/2023	22/12/2023
							F1905	Device Revision or Replacement	A0908	Incorrect, Inadequate or Imprecise Result or Readings	22/12/2023	22/12/2023
15864	2024	All other reportable incidents	37265	Dual-chamber implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
15865	2024	All other reportable incidents	16038	Implantable pulse generator mesh bag, non-bioabsorbable	CT1684	Cardiac pacemakers and associated devices	E210401	Shock from Patient Lead(s)	A071302	Inappropriate/Inadequate Shock/Stimulation		
							F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A071302	Inappropriate/Inadequate Shock/Stimulation		
							F2203	Imaging Required	A071302	Inappropriate/Inadequate Shock/Stimulation		
15866	2024	All other reportable incidents	47270	Cardiac resynchronization therapy implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A072001	Electromagnetic Interference		
									A090801	Signal Artifact/Noise		
							F26	No Health Consequences or Impact	A072001	Electromagnetic Interference		
									A090801	Signal Artifact/Noise		
15867	2024	Death	35852	Single-chamber implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E011903	Syncope/Fainting	A070101	Failure to Capture	02/09/2024	02/09/2024
									A0712	Pacing Problem	02/09/2024	02/09/2024
							E060101	Asystole	A070101	Failure to Capture	02/09/2024	02/09/2024
									A0712	Pacing Problem	02/09/2024	02/09/2024
							E233001	Chest Pain	A070101	Failure to Capture	02/09/2024	02/09/2024
									A0712	Pacing Problem	02/09/2024	02/09/2024
							F02	Death	A070101	Failure to Capture	02/09/2024	02/09/2024
									A0712	Pacing Problem	02/09/2024	02/09/2024
							F2306	Resuscitation	A070101	Failure to Capture	02/09/2024	02/09/2024
									A0712	Pacing Problem	02/09/2024	02/09/2024
							F2307	Reprogramming of Device	A070101	Failure to Capture	02/09/2024	02/09/2024
									A0712	Pacing Problem	02/09/2024	02/09/2024
15868	2024	All other reportable incidents	47265	Dual-chamber implantable pacemaker, rate-responsive	CT1685	Cardiac pacemakers	E012207	Twitching	A071209	Extracardiac Stimulation	30/08/2024	30/08/2024
									A0908	Incorrect, Inadequate or Imprecise Result or Readings	30/08/2024	30/08/2024
							E2103	Device Overstimulation of Tissue	A071209	Extracardiac Stimulation	30/08/2024	30/08/2024
									A0908	Incorrect, Inadequate or Imprecise Result or Readings	30/08/2024	30/08/2024
							F08	Hospitalization or Prolonged Hospitalization	A071209	Extracardiac Stimulation	30/08/2024	30/08/2024
									A0908	Incorrect, Inadequate or Imprecise Result or Readings	30/08/2024	30/08/2024
							F15	Recognized Device or Procedural Complication	A071209	Extracardiac Stimulation	30/08/2024	30/08/2024
									A0908	Incorrect, Inadequate or Imprecise Result or Readings	30/08/2024	30/08/2024

							F1905	Device Revision or Replacement	A071209	Extracardiac Stimulation	30/08/2024	30/08/2024
									A0908	Incorrect, Inadequate or Imprecise Result or Readings	30/08/2024	30/08/2024
					CT257	Implantable cardiac pacemakers	E012207	Twitching	A071209	Extracardiac Stimulation	30/08/2024	30/08/2024
									A0908	Incorrect, Inadequate or Imprecise Result or Readings	30/08/2024	30/08/2024
							E2103	Device Overstimulation of Tissue	A071209	Extracardiac Stimulation	30/08/2024	30/08/2024
									A0908	Incorrect, Inadequate or Imprecise Result or Readings	30/08/2024	30/08/2024
							F08	Hospitalization or Prolonged Hospitalization	A071209	Extracardiac Stimulation	30/08/2024	30/08/2024
									A0908	Incorrect, Inadequate or Imprecise Result or Readings	30/08/2024	30/08/2024
							F15	Recognized Device or Procedural Complication	A071209	Extracardiac Stimulation	30/08/2024	30/08/2024
									A0908	Incorrect, Inadequate or Imprecise Result or Readings	30/08/2024	30/08/2024
							F1905	Device Revision or Replacement	A071209	Extracardiac Stimulation	30/08/2024	30/08/2024
									A0908	Incorrect, Inadequate or Imprecise Result or Readings	30/08/2024	30/08/2024
					CT2866	Implantable cardiac electrical stimulators	E012207	Twitching	A071209	Extracardiac Stimulation	30/08/2024	30/08/2024
									A0908	Incorrect, Inadequate or Imprecise Result or Readings	30/08/2024	30/08/2024
							E2103	Device Overstimulation of Tissue	A071209	Extracardiac Stimulation	30/08/2024	30/08/2024
									A0908	Incorrect, Inadequate or Imprecise Result or Readings	30/08/2024	30/08/2024
							F08	Hospitalization or Prolonged Hospitalization	A071209	Extracardiac Stimulation	30/08/2024	30/08/2024
									A0908	Incorrect, Inadequate or Imprecise Result or Readings	30/08/2024	30/08/2024
							F15	Recognized Device or Procedural Complication	A071209	Extracardiac Stimulation	30/08/2024	30/08/2024
									A0908	Incorrect, Inadequate or Imprecise Result or Readings	30/08/2024	30/08/2024
							F1905	Device Revision or Replacement	A071209	Extracardiac Stimulation	30/08/2024	30/08/2024
									A0908	Incorrect, Inadequate or Imprecise Result or Readings	30/08/2024	30/08/2024
15869	2024	All other reportable incidents			CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A0203	Defective Device	28/08/2024	28/08/2024
									A070101	Failure to Capture	28/08/2024	28/08/2024
									A070102	High Capture Threshold	28/08/2024	28/08/2024
									A12	Connection Problem	28/08/2024	28/08/2024
							F08	Hospitalization or Prolonged Hospitalization	A0203	Defective Device	28/08/2024	28/08/2024
									A070101	Failure to Capture	28/08/2024	28/08/2024
									A070102	High Capture Threshold	28/08/2024	28/08/2024
									A12	Connection Problem	28/08/2024	28/08/2024
							F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A0203	Defective Device	28/08/2024	28/08/2024
									A070101	Failure to Capture	28/08/2024	28/08/2024
									A070102	High Capture Threshold	28/08/2024	28/08/2024
									A12	Connection Problem	28/08/2024	28/08/2024
							F1905	Device Revision or Replacement	A0203	Defective Device	28/08/2024	28/08/2024
									A070101	Failure to Capture	28/08/2024	28/08/2024
									A070102	High Capture Threshold	28/08/2024	28/08/2024
									A12	Connection Problem	28/08/2024	28/08/2024
							F2203	Imaging Required	A0203	Defective Device	28/08/2024	28/08/2024
									A070101	Failure to Capture	28/08/2024	28/08/2024
									A070102	High Capture Threshold	28/08/2024	28/08/2024
									A12	Connection Problem	28/08/2024	28/08/2024
15870	2024	All other reportable incidents	47265	Dual-chamber implantable pacemaker, rate-responsive	CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem	22/04/2024	22/04/2024
									A070504	Premature Discharge of Battery	22/04/2024	22/04/2024
							F08	Hospitalization or Prolonged Hospitalization	A0705	Battery Problem	22/04/2024	22/04/2024
									A070504	Premature Discharge of Battery	22/04/2024	22/04/2024
							F1905	Device Revision or Replacement	A0705	Battery Problem	22/04/2024	22/04/2024
									A070504	Premature Discharge of Battery	22/04/2024	22/04/2024
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem	22/04/2024	22/04/2024
									A070504	Premature Discharge of Battery	22/04/2024	22/04/2024
							F08	Hospitalization or Prolonged Hospitalization	A0705	Battery Problem	22/04/2024	22/04/2024
									A070504	Premature Discharge of Battery	22/04/2024	22/04/2024
							F1905	Device Revision or Replacement	A0705	Battery Problem	22/04/2024	22/04/2024
									A070504	Premature Discharge of Battery	22/04/2024	22/04/2024
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem	22/04/2024	22/04/2024
									A070504	Premature Discharge of Battery	22/04/2024	22/04/2024
							F08	Hospitalization or Prolonged Hospitalization	A0705	Battery Problem	22/04/2024	22/04/2024
									A070504	Premature Discharge of Battery	22/04/2024	22/04/2024
							F1905	Device Revision or Replacement	A0705	Battery Problem	22/04/2024	22/04/2024
									A070504	Premature Discharge of Battery	22/04/2024	22/04/2024
15871	2024	All other reportable incidents	35852	Single-chamber implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A070101	Failure to Capture		
									A070909	Oversensing		
							F26	No Health Consequences or Impact	A070101	Failure to Capture		
									A070909	Oversensing		
15872	2024	All other reportable incidents	35852	Single-chamber implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E0619	Pericardial Effusion	A090801	Signal Artifact/Noise		
							E2103	Device Overstimulation of Tissue	A090801	Signal Artifact/Noise		
							F08	Hospitalization or Prolonged Hospitalization	A090801	Signal Artifact/Noise		
15873	2024	All other reportable incidents	60956	Pacing/defibrillation lead tunneller	CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A0203	Defective Device		
									A070101	Failure to Capture		
							F08	Hospitalization or Prolonged Hospitalization	A0203	Defective Device		
									A070101	Failure to Capture		
							F1908	Prolonged Surgery	A0203	Defective Device		
									A070101	Failure to Capture		
15874	2024		47265	Dual-chamber implantable pacemaker, rate-responsive	CT141	Electrodes/Leads and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	31/05/2024	31/05/2024
							F1901	Additional Surgery	A24	No Apparent Device Problem	31/05/2024	31/05/2024
					CT145	Cardiovascular devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	31/05/2024	31/05/2024
							F1901	Additional Surgery	A24	No Apparent Device Problem	31/05/2024	31/05/2024
					CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	31/05/2024	31/05/2024
							F1901	Additional Surgery	A24	No Apparent Device Problem	31/05/2024	31/05/2024

					CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	31/05/2024	31/05/2024
							F1901	Additional Surgery	A24	No Apparent Device Problem	31/05/2024	31/05/2024
					CT233	Surgical	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	31/05/2024	31/05/2024
							F1901	Additional Surgery	A24	No Apparent Device Problem	31/05/2024	31/05/2024
					CT2406	Implantable	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	31/05/2024	31/05/2024
							F1901	Additional Surgery	A24	No Apparent Device Problem	31/05/2024	31/05/2024
					CT2407	Active implantable	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	31/05/2024	31/05/2024
							F1901	Additional Surgery	A24	No Apparent Device Problem	31/05/2024	31/05/2024
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	31/05/2024	31/05/2024
							F1901	Additional Surgery	A24	No Apparent Device Problem	31/05/2024	31/05/2024
					CT2727	Active implant devices/ancillaries	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	31/05/2024	31/05/2024
							F1901	Additional Surgery	A24	No Apparent Device Problem	31/05/2024	31/05/2024
					CT2750	Implantable cardiac electrical stimulators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	31/05/2024	31/05/2024
							F1901	Additional Surgery	A24	No Apparent Device Problem	31/05/2024	31/05/2024
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	31/05/2024	31/05/2024
							F1901	Additional Surgery	A24	No Apparent Device Problem	31/05/2024	31/05/2024
					CT2985	Active	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	31/05/2024	31/05/2024
							F1901	Additional Surgery	A24	No Apparent Device Problem	31/05/2024	31/05/2024
					CT3001	Implantable metal	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	31/05/2024	31/05/2024
							F1901	Additional Surgery	A24	No Apparent Device Problem	31/05/2024	31/05/2024
					CT3002	Implantable synthetic polymer	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	31/05/2024	31/05/2024
							F1901	Additional Surgery	A24	No Apparent Device Problem	31/05/2024	31/05/2024
					CT3006	Implantable inorganic material	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	31/05/2024	31/05/2024
							F1901	Additional Surgery	A24	No Apparent Device Problem	31/05/2024	31/05/2024
					CT314	Electrically powered	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	31/05/2024	31/05/2024
							F1901	Additional Surgery	A24	No Apparent Device Problem	31/05/2024	31/05/2024
					CT321	Long-term surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	31/05/2024	31/05/2024
							F1901	Additional Surgery	A24	No Apparent Device Problem	31/05/2024	31/05/2024
					CT449	Generators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	31/05/2024	31/05/2024
							F1901	Additional Surgery	A24	No Apparent Device Problem	31/05/2024	31/05/2024
					CT981	Single use	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	31/05/2024	31/05/2024
							F1901	Additional Surgery	A24	No Apparent Device Problem	31/05/2024	31/05/2024
					CT983	Surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	31/05/2024	31/05/2024
							F1901	Additional Surgery	A24	No Apparent Device Problem	31/05/2024	31/05/2024
15875	2024	All other reportable incidents	17882	Physiologic-monitoring defibrillation system	CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A0719	Unexpected Shutdown		
							F27	Problem Identified During Non-Clinical Procedure	A0719	Unexpected Shutdown		
15876	2024	All other reportable incidents	37265	Dual-chamber implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery		
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery		
15877	2024	All other reportable incidents	47265	Dual-chamber implantable pacemaker, rate-responsive	CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A13	Communication or Transmission Problem	06/09/2024	06/09/2024
							F26	No Health Consequences or Impact	A13	Communication or Transmission Problem	06/09/2024	06/09/2024
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A13	Communication or Transmission Problem	06/09/2024	06/09/2024
							F26	No Health Consequences or Impact	A13	Communication or Transmission Problem	06/09/2024	06/09/2024
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A13	Communication or Transmission Problem	06/09/2024	06/09/2024
							F26	No Health Consequences or Impact	A13	Communication or Transmission Problem	06/09/2024	06/09/2024
15878	2024	All other reportable incidents	35224	External pacemaker, epicardial pacing	CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A0719	Unexpected Shutdown		
									A190501	Moisture Damage		
							F23	Unexpected Medical Intervention	A0719	Unexpected Shutdown		
									A190501	Moisture Damage		
15879	2024	All other reportable incidents	35852	Single-chamber implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem	22/02/2024	22/02/2024
									A070504	Premature Discharge of Battery	22/02/2024	22/02/2024
							F08	Hospitalization or Prolonged Hospitalization	A0705	Battery Problem	22/02/2024	22/02/2024
									A070504	Premature Discharge of Battery	22/02/2024	22/02/2024
							F1905	Device Revision or Replacement	A0705	Battery Problem	22/02/2024	22/02/2024
									A070504	Premature Discharge of Battery	22/02/2024	22/02/2024
15880	2024	All other reportable incidents	47265	Dual-chamber implantable pacemaker, rate-responsive	CT141	Electrodes/Leads and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A0908	Incorrect, Inadequate or Imprecise Result or Readings	16/02/2024	16/02/2024
							F08	Hospitalization or Prolonged Hospitalization	A0908	Incorrect, Inadequate or Imprecise Result or Readings	16/02/2024	16/02/2024
							F1903	Device Explantation	A0908	Incorrect, Inadequate or Imprecise Result or Readings	16/02/2024	16/02/2024
					CT145	Cardiovascular devices	E2403	No Clinical Signs, Symptoms or Conditions	A0908	Incorrect, Inadequate or Imprecise Result or Readings	16/02/2024	16/02/2024
							F08	Hospitalization or Prolonged Hospitalization	A0908	Incorrect, Inadequate or Imprecise Result or Readings	16/02/2024	16/02/2024
							F1903	Device Explantation	A0908	Incorrect, Inadequate or Imprecise Result or Readings	16/02/2024	16/02/2024
					CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A0908	Incorrect, Inadequate or Imprecise Result or Readings	16/02/2024	16/02/2024
							F08	Hospitalization or Prolonged Hospitalization	A0908	Incorrect, Inadequate or Imprecise Result or Readings	16/02/2024	16/02/2024
							F1903	Device Explantation	A0908	Incorrect, Inadequate or Imprecise Result or Readings	16/02/2024	16/02/2024
					CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A0908	Incorrect, Inadequate or Imprecise Result or Readings	16/02/2024	16/02/2024
							F08	Hospitalization or Prolonged Hospitalization	A0908	Incorrect, Inadequate or Imprecise Result or Readings	16/02/2024	16/02/2024
							F1903	Device Explantation	A0908	Incorrect, Inadequate or Imprecise Result or Readings	16/02/2024	16/02/2024
					CT233	Surgical	E2403	No Clinical Signs, Symptoms or Conditions	A0908	Incorrect, Inadequate or Imprecise Result or Readings	16/02/2024	16/02/2024
							F08	Hospitalization or Prolonged Hospitalization	A0908	Incorrect, Inadequate or Imprecise Result or Readings	16/02/2024	16/02/2024
							F1903	Device Explantation	A0908	Incorrect, Inadequate or Imprecise Result or Readings	16/02/2024	16/02/2024
					CT2406	Implantable	E2403	No Clinical Signs, Symptoms or Conditions	A0908	Incorrect, Inadequate or Imprecise Result or Readings	16/02/2024	16/02/2024
							F08	Hospitalization or Prolonged Hospitalization	A0908	Incorrect, Inadequate or Imprecise Result or Readings	16/02/2024	16/02/2024
							F1903	Device Explantation	A0908	Incorrect, Inadequate or Imprecise Result or Readings	16/02/2024	16/02/2024

					CT2407	Active implantable	E2403	No Clinical Signs, Symptoms or Conditions	A0908	Incorrect, Inadequate or Imprecise Result or Readings	16/02/2024	16/02/2024
							F08	Hospitalization or Prolonged Hospitalization	A0908	Incorrect, Inadequate or Imprecise Result or Readings	16/02/2024	16/02/2024
							F1903	Device Explantation	A0908	Incorrect, Inadequate or Imprecise Result or Readings	16/02/2024	16/02/2024
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A0908	Incorrect, Inadequate or Imprecise Result or Readings	16/02/2024	16/02/2024
							F08	Hospitalization or Prolonged Hospitalization	A0908	Incorrect, Inadequate or Imprecise Result or Readings	16/02/2024	16/02/2024
							F1903	Device Explantation	A0908	Incorrect, Inadequate or Imprecise Result or Readings	16/02/2024	16/02/2024
					CT2727	Active implant devices/ancillaries	E2403	No Clinical Signs, Symptoms or Conditions	A0908	Incorrect, Inadequate or Imprecise Result or Readings	16/02/2024	16/02/2024
							F08	Hospitalization or Prolonged Hospitalization	A0908	Incorrect, Inadequate or Imprecise Result or Readings	16/02/2024	16/02/2024
							F1903	Device Explantation	A0908	Incorrect, Inadequate or Imprecise Result or Readings	16/02/2024	16/02/2024
					CT2750	Implantable cardiac electrical stimulators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A0908	Incorrect, Inadequate or Imprecise Result or Readings	16/02/2024	16/02/2024
							F08	Hospitalization or Prolonged Hospitalization	A0908	Incorrect, Inadequate or Imprecise Result or Readings	16/02/2024	16/02/2024
							F1903	Device Explantation	A0908	Incorrect, Inadequate or Imprecise Result or Readings	16/02/2024	16/02/2024
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A0908	Incorrect, Inadequate or Imprecise Result or Readings	16/02/2024	16/02/2024
							F08	Hospitalization or Prolonged Hospitalization	A0908	Incorrect, Inadequate or Imprecise Result or Readings	16/02/2024	16/02/2024
							F1903	Device Explantation	A0908	Incorrect, Inadequate or Imprecise Result or Readings	16/02/2024	16/02/2024
					CT2985	Active	E2403	No Clinical Signs, Symptoms or Conditions	A0908	Incorrect, Inadequate or Imprecise Result or Readings	16/02/2024	16/02/2024
							F08	Hospitalization or Prolonged Hospitalization	A0908	Incorrect, Inadequate or Imprecise Result or Readings	16/02/2024	16/02/2024
							F1903	Device Explantation	A0908	Incorrect, Inadequate or Imprecise Result or Readings	16/02/2024	16/02/2024
					CT3001	Implantable metal	E2403	No Clinical Signs, Symptoms or Conditions	A0908	Incorrect, Inadequate or Imprecise Result or Readings	16/02/2024	16/02/2024
							F08	Hospitalization or Prolonged Hospitalization	A0908	Incorrect, Inadequate or Imprecise Result or Readings	16/02/2024	16/02/2024
							F1903	Device Explantation	A0908	Incorrect, Inadequate or Imprecise Result or Readings	16/02/2024	16/02/2024
					CT3002	Implantable synthetic polymer	E2403	No Clinical Signs, Symptoms or Conditions	A0908	Incorrect, Inadequate or Imprecise Result or Readings	16/02/2024	16/02/2024
							F08	Hospitalization or Prolonged Hospitalization	A0908	Incorrect, Inadequate or Imprecise Result or Readings	16/02/2024	16/02/2024
							F1903	Device Explantation	A0908	Incorrect, Inadequate or Imprecise Result or Readings	16/02/2024	16/02/2024
					CT3006	Implantable inorganic material	E2403	No Clinical Signs, Symptoms or Conditions	A0908	Incorrect, Inadequate or Imprecise Result or Readings	16/02/2024	16/02/2024
							F08	Hospitalization or Prolonged Hospitalization	A0908	Incorrect, Inadequate or Imprecise Result or Readings	16/02/2024	16/02/2024
							F1903	Device Explantation	A0908	Incorrect, Inadequate or Imprecise Result or Readings	16/02/2024	16/02/2024
					CT314	Electrically powered	E2403	No Clinical Signs, Symptoms or Conditions	A0908	Incorrect, Inadequate or Imprecise Result or Readings	16/02/2024	16/02/2024
							F08	Hospitalization or Prolonged Hospitalization	A0908	Incorrect, Inadequate or Imprecise Result or Readings	16/02/2024	16/02/2024
							F1903	Device Explantation	A0908	Incorrect, Inadequate or Imprecise Result or Readings	16/02/2024	16/02/2024
					CT321	Long-term surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A0908	Incorrect, Inadequate or Imprecise Result or Readings	16/02/2024	16/02/2024
							F08	Hospitalization or Prolonged Hospitalization	A0908	Incorrect, Inadequate or Imprecise Result or Readings	16/02/2024	16/02/2024
							F1903	Device Explantation	A0908	Incorrect, Inadequate or Imprecise Result or Readings	16/02/2024	16/02/2024
					CT449	Generators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A0908	Incorrect, Inadequate or Imprecise Result or Readings	16/02/2024	16/02/2024
							F08	Hospitalization or Prolonged Hospitalization	A0908	Incorrect, Inadequate or Imprecise Result or Readings	16/02/2024	16/02/2024
							F1903	Device Explantation	A0908	Incorrect, Inadequate or Imprecise Result or Readings	16/02/2024	16/02/2024
					CT981	Single use	E2403	No Clinical Signs, Symptoms or Conditions	A0908	Incorrect, Inadequate or Imprecise Result or Readings	16/02/2024	16/02/2024
							F08	Hospitalization or Prolonged Hospitalization	A0908	Incorrect, Inadequate or Imprecise Result or Readings	16/02/2024	16/02/2024
							F1903	Device Explantation	A0908	Incorrect, Inadequate or Imprecise Result or Readings	16/02/2024	16/02/2024
					CT983	Surgical invasive	E2403	No Clinical Signs, Symptoms or Conditions	A0908	Incorrect, Inadequate or Imprecise Result or Readings	16/02/2024	16/02/2024
							F08	Hospitalization or Prolonged Hospitalization	A0908	Incorrect, Inadequate or Imprecise Result or Readings	16/02/2024	16/02/2024
							F1903	Device Explantation	A0908	Incorrect, Inadequate or Imprecise Result or Readings	16/02/2024	16/02/2024
15881	2024	All other reportable incidents	47263	Cardiac resynchronization therapy implantable pacemaker	CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A0908	Incorrect, Inadequate or Imprecise Result or Readings	03/09/2024	03/09/2024
							F08	Hospitalization or Prolonged Hospitalization	A0908	Incorrect, Inadequate or Imprecise Result or Readings	03/09/2024	03/09/2024
							F1905	Device Revision or Replacement	A0908	Incorrect, Inadequate or Imprecise Result or Readings	03/09/2024	03/09/2024
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A0908	Incorrect, Inadequate or Imprecise Result or Readings	03/09/2024	03/09/2024
							F08	Hospitalization or Prolonged Hospitalization	A0908	Incorrect, Inadequate or Imprecise Result or Readings	03/09/2024	03/09/2024
							F1905	Device Revision or Replacement	A0908	Incorrect, Inadequate or Imprecise Result or Readings	03/09/2024	03/09/2024
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A0908	Incorrect, Inadequate or Imprecise Result or Readings	03/09/2024	03/09/2024
							F08	Hospitalization or Prolonged Hospitalization	A0908	Incorrect, Inadequate or Imprecise Result or Readings	03/09/2024	03/09/2024

							F1905	Device Revision or Replacement	A0908	Incorrect, Inadequate or Imprecise Result or Readings	03/09/2024	03/09/2024
15882	2024	All other reportable incidents	47270	Cardiac resynchronization therapy implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A072202	Low Impedance		
							F2203	Imaging Required	A072202	Low Impedance		
15883	2024	All other reportable incidents	17882	Physiologic-monitoring defibrillation system	CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A0801	Failure to Calibrate		
							F27	Problem Identified During Non-Clinical Procedure	A0801	Failure to Calibrate		
15884	2024	All other reportable incidents	17882	Physiologic-monitoring defibrillation system	CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A0801	Failure to Calibrate		
							F27	Problem Identified During Non-Clinical Procedure	A0801	Failure to Calibrate		
15885	2024	All other reportable incidents	17882	Physiologic-monitoring defibrillation system	CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A0801	Failure to Calibrate		
							F27	Problem Identified During Non-Clinical Procedure	A0801	Failure to Calibrate		
15886	2024	All other reportable incidents	17882	Physiologic-monitoring defibrillation system	CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A0801	Failure to Calibrate		
							F27	Problem Identified During Non-Clinical Procedure	A0801	Failure to Calibrate		
15887	2024	All other reportable incidents	17882	Physiologic-monitoring defibrillation system	CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A071301	Failure to Deliver Shock/Stimulation		
							F26	No Health Consequences or Impact	A071301	Failure to Deliver Shock/Stimulation		
15888	2024		47263	Cardiac resynchronization therapy implantable pacemaker	CT1685	Cardiac pacemakers	F26	No Health Consequences or Impact	A02	Manufacturing, Packaging or Shipping Problem		
					CT257	Implantable cardiac pacemakers	F26	No Health Consequences or Impact	A02	Manufacturing, Packaging or Shipping Problem		
					CT2866	Implantable cardiac electrical stimulators	F26	No Health Consequences or Impact	A02	Manufacturing, Packaging or Shipping Problem		
15889	2024	All other reportable incidents	35852	Single-chamber implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	14/05/2024	14/05/2024
									A071102	Failure to Interrogate	14/05/2024	14/05/2024
							F1901	Additional Surgery	A070504	Premature Discharge of Battery	14/05/2024	14/05/2024
									A071102	Failure to Interrogate	14/05/2024	14/05/2024
15890	2024	All other reportable incidents	47265	Dual-chamber implantable pacemaker, rate-responsive	CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A0712	Pacing Problem	23/06/2024	23/06/2024
							F08	Hospitalization or Prolonged Hospitalization	A0712	Pacing Problem	23/06/2024	23/06/2024
							F12	Serious Injury/ Illness/ Impairment	A0712	Pacing Problem	23/06/2024	23/06/2024
							F1905	Device Revision or Replacement	A0712	Pacing Problem	23/06/2024	23/06/2024
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A0712	Pacing Problem	23/06/2024	23/06/2024
							F08	Hospitalization or Prolonged Hospitalization	A0712	Pacing Problem	23/06/2024	23/06/2024
							F12	Serious Injury/ Illness/ Impairment	A0712	Pacing Problem	23/06/2024	23/06/2024
							F1905	Device Revision or Replacement	A0712	Pacing Problem	23/06/2024	23/06/2024
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A0712	Pacing Problem	23/06/2024	23/06/2024
							F08	Hospitalization or Prolonged Hospitalization	A0712	Pacing Problem	23/06/2024	23/06/2024
							F12	Serious Injury/ Illness/ Impairment	A0712	Pacing Problem	23/06/2024	23/06/2024
							F1905	Device Revision or Replacement	A0712	Pacing Problem	23/06/2024	23/06/2024
15891	2024	All other reportable incidents	17882	Physiologic-monitoring defibrillation system	CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A07	Electrical /Electronic Property Problem		
									A0719	Unexpected Shutdown		
							F26	No Health Consequences or Impact	A07	Electrical /Electronic Property Problem		
									A0719	Unexpected Shutdown		
15892	2024	All other reportable incidents	35852	Single-chamber implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	13/02/2024	13/02/2024
									A0908	Incorrect, Inadequate or Imprecise Result or Readings	13/02/2024	13/02/2024
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	13/02/2024	13/02/2024
									A0908	Incorrect, Inadequate or Imprecise Result or Readings	13/02/2024	13/02/2024
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	13/02/2024	13/02/2024
									A0908	Incorrect, Inadequate or Imprecise Result or Readings	13/02/2024	13/02/2024
15893	2024	All other reportable incidents	47270	Cardiac resynchronization therapy implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A070909	Oversensing		
							F1901	Additional Surgery	A070909	Oversensing		
15894	2024	All other reportable incidents	47270	Cardiac resynchronization therapy implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A070909	Oversensing	07/06/2024	07/06/2024
							F1901	Additional Surgery	A070909	Oversensing	07/06/2024	07/06/2024
15895	2024	All other reportable incidents			CT1684	Cardiac pacemakers and associated devices	E2103	Device Overstimulation of Tissue	A070101	Failure to Capture	06/11/2024	06/11/2024
									A070102	High Capture Threshold	06/11/2024	06/11/2024
							F08	Hospitalization or Prolonged Hospitalization	A070101	Failure to Capture	06/11/2024	06/11/2024
									A070102	High Capture Threshold	06/11/2024	06/11/2024
							F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A070101	Failure to Capture	06/11/2024	06/11/2024
									A070102	High Capture Threshold	06/11/2024	06/11/2024
							F1905	Device Revision or Replacement	A070101	Failure to Capture	06/11/2024	06/11/2024
									A070102	High Capture Threshold	06/11/2024	06/11/2024
15896	2024	All other reportable incidents	47265	Dual-chamber implantable pacemaker, rate-responsive	CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	28/08/2024	28/08/2024
							F1901	Additional Surgery	A24	No Apparent Device Problem	28/08/2024	28/08/2024
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	28/08/2024	28/08/2024
							F1901	Additional Surgery	A24	No Apparent Device Problem	28/08/2024	28/08/2024
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	28/08/2024	28/08/2024
							F1901	Additional Surgery	A24	No Apparent Device Problem	28/08/2024	28/08/2024
15897	2024	All other reportable incidents	35852	Single-chamber implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem	18/07/2024	18/07/2024
									A070504	Premature Discharge of Battery	18/07/2024	18/07/2024
							F08	Hospitalization or Prolonged Hospitalization	A0705	Battery Problem	18/07/2024	18/07/2024
									A070504	Premature Discharge of Battery	18/07/2024	18/07/2024
							F1905	Device Revision or Replacement	A0705	Battery Problem	18/07/2024	18/07/2024
									A070504	Premature Discharge of Battery	18/07/2024	18/07/2024
15898	2024	All other reportable incidents	35852	Single-chamber implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A0908	Incorrect, Inadequate or Imprecise Result or Readings		
							F27	Problem Identified During Non-Clinical Procedure	A0908	Incorrect, Inadequate or Imprecise Result or Readings		
15899	2024		37265	Dual-chamber implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery		
15900	2024	All other reportable incidents	37265	Dual-chamber implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A160602	Premature End-of-Life Indicator	23/02/2024	23/02/2024
							F1905	Device Revision or Replacement	A160602	Premature End-of-Life Indicator	23/02/2024	23/02/2024
15901	2024	All other reportable incidents	47205	Cardiac pulse generator programmer	CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A0902	Display or Visual Feedback Problem		
							F27	Problem Identified During Non-Clinical Procedure	A0902	Display or Visual Feedback Problem		
15902	2024	All other reportable incidents	47265	Dual-chamber implantable pacemaker, rate-responsive	CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A0908	Incorrect, Inadequate or Imprecise Result or Readings		
							F08	Hospitalization or Prolonged Hospitalization	A0908	Incorrect, Inadequate or Imprecise Result or Readings		
							F1905	Device Revision or Replacement	A0908	Incorrect, Inadequate or Imprecise Result or Readings		
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A0908	Incorrect, Inadequate or Imprecise Result or Readings		
							F08	Hospitalization or Prolonged Hospitalization	A0908	Incorrect, Inadequate or Imprecise Result or Readings		

							F1905	Device Revision or Replacement	A0908	Incorrect, Inadequate or Imprecise Result or Readings			
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A0908	Incorrect, Inadequate or Imprecise Result or Readings			
							F08	Hospitalization or Prolonged Hospitalization	A0908	Incorrect, Inadequate or Imprecise Result or Readings			
							F1905	Device Revision or Replacement	A0908	Incorrect, Inadequate or Imprecise Result or Readings			
15903	2024	All other reportable incidents	35852	Single-chamber implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery			
									A0908	Incorrect, Inadequate or Imprecise Result or Readings			
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery			
									A0908	Incorrect, Inadequate or Imprecise Result or Readings			
15904	2024	All other reportable incidents	47265	Dual-chamber implantable pacemaker, rate-responsive	CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem			
									A070504	Premature Discharge of Battery			
							F26	No Health Consequences or Impact	A0705	Battery Problem			
									A070504	Premature Discharge of Battery			
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem			
									A070504	Premature Discharge of Battery			
							F26	No Health Consequences or Impact	A0705	Battery Problem			
									A070504	Premature Discharge of Battery			
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem			
									A070504	Premature Discharge of Battery			
							F26	No Health Consequences or Impact	A0705	Battery Problem			
									A070504	Premature Discharge of Battery			
15905	2024	All other reportable incidents	47270	Cardiac resynchronization therapy implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	18/10/2024	18/10/2024	
									A0908	Incorrect, Inadequate or Imprecise Result or Readings	18/10/2024	18/10/2024	
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	18/10/2024	18/10/2024	
									A0908	Incorrect, Inadequate or Imprecise Result or Readings	18/10/2024	18/10/2024	
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	18/10/2024	18/10/2024	
									A0908	Incorrect, Inadequate or Imprecise Result or Readings	18/10/2024	18/10/2024	
15906	2024	All other reportable incidents	47267	Single-chamber implantable pacemaker, rate-responsive	CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A05	Mechanical Problem	06/09/2024	06/09/2024	
									A0712	Pacing Problem	06/09/2024	06/09/2024	
							F26	No Health Consequences or Impact	A05	Mechanical Problem	06/09/2024	06/09/2024	
									A0712	Pacing Problem	06/09/2024	06/09/2024	
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A05	Mechanical Problem	06/09/2024	06/09/2024	
									A0712	Pacing Problem	06/09/2024	06/09/2024	
							F26	No Health Consequences or Impact	A05	Mechanical Problem	06/09/2024	06/09/2024	
									A0712	Pacing Problem	06/09/2024	06/09/2024	
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A05	Mechanical Problem	06/09/2024	06/09/2024	
									A0712	Pacing Problem	06/09/2024	06/09/2024	
							F26	No Health Consequences or Impact	A05	Mechanical Problem	06/09/2024	06/09/2024	
									A0712	Pacing Problem	06/09/2024	06/09/2024	
15907	2024	All other reportable incidents	35852	Single-chamber implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E210401	Shock from Patient Lead(s)	A0709	Device Sensing Problem			
									A071302	Inappropriate/Inadequate Shock/Stimulation			
							F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A0709	Device Sensing Problem			
									A071302	Inappropriate/Inadequate Shock/Stimulation			
							F1203	Life Threatening Illness or Injury	A0709	Device Sensing Problem			
									A071302	Inappropriate/Inadequate Shock/Stimulation			
							F2303	Medication Required	A0709	Device Sensing Problem			
									A071302	Inappropriate/Inadequate Shock/Stimulation			
15908	2024	Death	17882	Physiologic-monitoring defibrillation system	CT1684	Cardiac pacemakers and associated devices			A0707	Failure to Discharge			
15909	2024	All other reportable incidents	35852	Single-chamber implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A0508	Noise, Audible	22/09/2024	22/09/2024	
									A070504	Premature Discharge of Battery	22/09/2024	22/09/2024	
									A13	Communication or Transmission Problem	22/09/2024	22/09/2024	
							F08	Hospitalization or Prolonged Hospitalization	A0508	Noise, Audible	22/09/2024	22/09/2024	
									A070504	Premature Discharge of Battery	22/09/2024	22/09/2024	
									A13	Communication or Transmission Problem	22/09/2024	22/09/2024	
							F1905	Device Revision or Replacement	A0508	Noise, Audible	22/09/2024	22/09/2024	
									A070504	Premature Discharge of Battery	22/09/2024	22/09/2024	
									A13	Communication or Transmission Problem	22/09/2024	22/09/2024	
15910	2024	All other reportable incidents	35852	Single-chamber implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A160601	Premature Elective Replacement Indicator			
							F1903	Device Explantation	A160601	Premature Elective Replacement Indicator			
15911	2024		47265	Dual-chamber implantable pacemaker, rate-responsive	CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	10/09/2024	10/09/2024	
							F1901	Additional Surgery	A24	No Apparent Device Problem	10/09/2024	10/09/2024	
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	10/09/2024	10/09/2024	
							F1901	Additional Surgery	A24	No Apparent Device Problem	10/09/2024	10/09/2024	
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	10/09/2024	10/09/2024	
							F1901	Additional Surgery	A24	No Apparent Device Problem	10/09/2024	10/09/2024	
15912	2024	All other reportable incidents	37265	Dual-chamber implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A072201	High Impedance			
							F26	No Health Consequences or Impact	A072201	High Impedance			
15913	2024	All other reportable incidents	47265	Dual-chamber implantable pacemaker, rate-responsive	CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A12	Connection Problem			
							F1908	Prolonged Surgery	A12	Connection Problem			
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A12	Connection Problem			
							F1908	Prolonged Surgery	A12	Connection Problem			
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A12	Connection Problem			
							F1908	Prolonged Surgery	A12	Connection Problem			
15914	2024		47265	Dual-chamber implantable pacemaker, rate-responsive	CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A070101	Failure to Capture			
									A071206	Pacing Asynchronously			
							F26	No Health Consequences or Impact	A070101	Failure to Capture			
									A071206	Pacing Asynchronously			
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A070101	Failure to Capture			
									A071206	Pacing Asynchronously			
							F26	No Health Consequences or Impact	A070101	Failure to Capture			
									A071206	Pacing Asynchronously			
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A070101	Failure to Capture			

									A071206	Pacing Asynchronously			
							F26	No Health Consequences or Impact	A070101	Failure to Capture			
									A071206	Pacing Asynchronously			
15915	2024	All other reportable incidents			CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A12	Connection Problem			
							F1908	Prolonged Surgery	A12	Connection Problem			
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A12	Connection Problem			
							F1908	Prolonged Surgery	A12	Connection Problem			
15916	2024	All other reportable incidents	47270	Cardiac resynchronization therapy implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	18/02/2025	18/02/2025	
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	18/02/2025	18/02/2025	
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	18/02/2025	18/02/2025	
15917	2024	All other reportable incidents	17882	Physiologic-monitoring defibrillation system	CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A071301	Failure to Deliver Shock/Stimulation			
							F27	Problem Identified During Non-Clinical Procedure	A071301	Failure to Deliver Shock/Stimulation			
15918	2024	All other reportable incidents	17882	Physiologic-monitoring defibrillation system	CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A0801	Failure to Calibrate			
							F27	Problem Identified During Non-Clinical Procedure	A0801	Failure to Calibrate			
15919	2024	All other reportable incidents	17882	Physiologic-monitoring defibrillation system	CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A0908	Incorrect, Inadequate or Imprecise Result or Readings			
							F27	Problem Identified During Non-Clinical Procedure	A0908	Incorrect, Inadequate or Imprecise Result or Readings			
15920	2024	All other reportable incidents	17882	Physiologic-monitoring defibrillation system	CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070912	Incorrect Interpretation of Signal			
							F26	No Health Consequences or Impact	A070912	Incorrect Interpretation of Signal			
15921	2024	All other reportable incidents	47265	Dual-chamber implantable pacemaker, rate-responsive	CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem			
									A070504	Premature Discharge of Battery			
							F26	No Health Consequences or Impact	A0705	Battery Problem			
									A070504	Premature Discharge of Battery			
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem			
									A070504	Premature Discharge of Battery			
							F26	No Health Consequences or Impact	A0705	Battery Problem			
									A070504	Premature Discharge of Battery			
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem			
									A070504	Premature Discharge of Battery			
							F26	No Health Consequences or Impact	A0705	Battery Problem			
									A070504	Premature Discharge of Battery			
15922	2024	All other reportable incidents	35852	Single-chamber implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A1605	Reset Problem			
							F12	Serious Injury/ Illness/ Impairment	A1605	Reset Problem			
							F2307	Reprogramming of Device	A1605	Reset Problem			
15923	2024	All other reportable incidents	47265	Dual-chamber implantable pacemaker, rate-responsive	CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	05/08/2024	05/08/2024	
							F1901	Additional Surgery	A24	No Apparent Device Problem	05/08/2024	05/08/2024	
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	05/08/2024	05/08/2024	
							F1901	Additional Surgery	A24	No Apparent Device Problem	05/08/2024	05/08/2024	
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	05/08/2024	05/08/2024	
							F1901	Additional Surgery	A24	No Apparent Device Problem	05/08/2024	05/08/2024	
15924	2024		47265	Dual-chamber implantable pacemaker, rate-responsive	CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	13/09/2024	13/09/2024	
							F1901	Additional Surgery	A24	No Apparent Device Problem	13/09/2024	13/09/2024	
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	13/09/2024	13/09/2024	
							F1901	Additional Surgery	A24	No Apparent Device Problem	13/09/2024	13/09/2024	
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	13/09/2024	13/09/2024	
							F1901	Additional Surgery	A24	No Apparent Device Problem	13/09/2024	13/09/2024	
15925	2024	All other reportable incidents	47270	Cardiac resynchronization therapy implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E060110	Ventricular Fibrillation	A070602	Delayed Charge Time	13/01/2025	13/01/2025	
									A0711	Interrogation Problem	13/01/2025	13/01/2025	
									A071201	Failure to Convert Rhythm	13/01/2025	13/01/2025	
									A072202	Low Impedance	13/01/2025	13/01/2025	
							E2311	Discomfort	A070602	Delayed Charge Time	13/01/2025	13/01/2025	
									A0711	Interrogation Problem	13/01/2025	13/01/2025	
									A071201	Failure to Convert Rhythm	13/01/2025	13/01/2025	
									A072202	Low Impedance	13/01/2025	13/01/2025	
							E2340	Wound Dehiscence	A070602	Delayed Charge Time	13/01/2025	13/01/2025	
									A0711	Interrogation Problem	13/01/2025	13/01/2025	
									A071201	Failure to Convert Rhythm	13/01/2025	13/01/2025	
									A072202	Low Impedance	13/01/2025	13/01/2025	
							F08	Hospitalization or Prolonged Hospitalization	A070602	Delayed Charge Time	13/01/2025	13/01/2025	
									A0711	Interrogation Problem	13/01/2025	13/01/2025	
									A071201	Failure to Convert Rhythm	13/01/2025	13/01/2025	
									A072202	Low Impedance	13/01/2025	13/01/2025	
							F1203	Life Threatening Illness or Injury	A070602	Delayed Charge Time	13/01/2025	13/01/2025	
									A0711	Interrogation Problem	13/01/2025	13/01/2025	
									A071201	Failure to Convert Rhythm	13/01/2025	13/01/2025	
									A072202	Low Impedance	13/01/2025	13/01/2025	
							F1905	Device Revision or Replacement	A070602	Delayed Charge Time	13/01/2025	13/01/2025	
									A0711	Interrogation Problem	13/01/2025	13/01/2025	
									A071201	Failure to Convert Rhythm	13/01/2025	13/01/2025	
									A072202	Low Impedance	13/01/2025	13/01/2025	
							F23	Unexpected Medical Intervention	A070602	Delayed Charge Time	13/01/2025	13/01/2025	
									A0711	Interrogation Problem	13/01/2025	13/01/2025	
									A071201	Failure to Convert Rhythm	13/01/2025	13/01/2025	
									A072202	Low Impedance	13/01/2025	13/01/2025	
							F2306	Resuscitation	A070602	Delayed Charge Time	13/01/2025	13/01/2025	
									A0711	Interrogation Problem	13/01/2025	13/01/2025	
									A071201	Failure to Convert Rhythm	13/01/2025	13/01/2025	
									A072202	Low Impedance	13/01/2025	13/01/2025	
15926	2024	All other reportable incidents	17882	Physiologic-monitoring defibrillation system	CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A09	Output Problem			
							F27	Problem Identified During Non-Clinical Procedure	A09	Output Problem			
15927	2024	All other reportable incidents	17882	Physiologic-monitoring defibrillation system	CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A090201	Device Displays Incorrect Message			
							F27	Problem Identified During Non-Clinical Procedure	A090201	Device Displays Incorrect Message			
15928	2024	All other reportable incidents	17882	Physiologic-monitoring defibrillation system	CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A0801	Failure to Calibrate			
							F27	Problem Identified During Non-Clinical Procedure	A0801	Failure to Calibrate			
15929	2024	All other reportable incidents	47265	Dual-chamber implantable pacemaker, rate-responsive	CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	29/05/2024	29/05/2024	
							F1901	Additional Surgery	A24	No Apparent Device Problem	29/05/2024	29/05/2024	
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	29/05/2024	29/05/2024	

							F1901	Additional Surgery	A24	No Apparent Device Problem	29/05/2024	29/05/2024
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	29/05/2024	29/05/2024
							F1901	Additional Surgery	A24	No Apparent Device Problem	29/05/2024	29/05/2024
15930	2024		47265	Dual-chamber implantable pacemaker, rate-responsive	CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	31/05/2024	31/05/2024
							F1901	Additional Surgery	A24	No Apparent Device Problem	31/05/2024	31/05/2024
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	31/05/2024	31/05/2024
							F1901	Additional Surgery	A24	No Apparent Device Problem	31/05/2024	31/05/2024
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	31/05/2024	31/05/2024
							F1901	Additional Surgery	A24	No Apparent Device Problem	31/05/2024	31/05/2024
15931	2024	All other reportable incidents	47265	Dual-chamber implantable pacemaker, rate-responsive	CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A05	Mechanical Problem		
							F26	No Health Consequences or Impact	A05	Mechanical Problem		
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A05	Mechanical Problem		
							F26	No Health Consequences or Impact	A05	Mechanical Problem		
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A05	Mechanical Problem		
							F26	No Health Consequences or Impact	A05	Mechanical Problem		
15932	2024	All other reportable incidents	35852	Single-chamber implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A0709	Device Sensing Problem		
									A070909	Oversensing		
									A072001	Electromagnetic Interference		
									A090801	Signal Artifact/Noise		
							F12	Serious Injury/ Illness/ Impairment	A0709	Device Sensing Problem		
									A070909	Oversensing		
									A072001	Electromagnetic Interference		
									A090801	Signal Artifact/Noise		
							F2307	Reprogramming of Device	A0709	Device Sensing Problem		
									A070909	Oversensing		
									A072001	Electromagnetic Interference		
									A090801	Signal Artifact/Noise		
15933	2024	All other reportable incidents	47263	Cardiac resynchronization therapy implantable pacemaker	CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A0908	Incorrect, Inadequate or Imprecise Result or Readings	13/09/2024	13/09/2024
							F08	Hospitalization or Prolonged Hospitalization	A0908	Incorrect, Inadequate or Imprecise Result or Readings	13/09/2024	13/09/2024
							F1905	Device Revision or Replacement	A0908	Incorrect, Inadequate or Imprecise Result or Readings	13/09/2024	13/09/2024
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A0908	Incorrect, Inadequate or Imprecise Result or Readings	13/09/2024	13/09/2024
							F08	Hospitalization or Prolonged Hospitalization	A0908	Incorrect, Inadequate or Imprecise Result or Readings	13/09/2024	13/09/2024
							F1905	Device Revision or Replacement	A0908	Incorrect, Inadequate or Imprecise Result or Readings	13/09/2024	13/09/2024
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A0908	Incorrect, Inadequate or Imprecise Result or Readings	13/09/2024	13/09/2024
							F08	Hospitalization or Prolonged Hospitalization	A0908	Incorrect, Inadequate or Imprecise Result or Readings	13/09/2024	13/09/2024
							F1905	Device Revision or Replacement	A0908	Incorrect, Inadequate or Imprecise Result or Readings	13/09/2024	13/09/2024
15934	2024	All other reportable incidents	47265	Dual-chamber implantable pacemaker, rate-responsive	CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A0908	Incorrect, Inadequate or Imprecise Result or Readings	06/09/2024	06/09/2024
							F08	Hospitalization or Prolonged Hospitalization	A0908	Incorrect, Inadequate or Imprecise Result or Readings	06/09/2024	06/09/2024
							F1905	Device Revision or Replacement	A0908	Incorrect, Inadequate or Imprecise Result or Readings	06/09/2024	06/09/2024
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A0908	Incorrect, Inadequate or Imprecise Result or Readings	06/09/2024	06/09/2024
							F08	Hospitalization or Prolonged Hospitalization	A0908	Incorrect, Inadequate or Imprecise Result or Readings	06/09/2024	06/09/2024
							F1905	Device Revision or Replacement	A0908	Incorrect, Inadequate or Imprecise Result or Readings	06/09/2024	06/09/2024
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A0908	Incorrect, Inadequate or Imprecise Result or Readings	06/09/2024	06/09/2024
							F08	Hospitalization or Prolonged Hospitalization	A0908	Incorrect, Inadequate or Imprecise Result or Readings	06/09/2024	06/09/2024
							F1905	Device Revision or Replacement	A0908	Incorrect, Inadequate or Imprecise Result or Readings	06/09/2024	06/09/2024
15935	2024		47265	Dual-chamber implantable pacemaker, rate-responsive	CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	30/06/2024	30/06/2024
							F1901	Additional Surgery	A24	No Apparent Device Problem	30/06/2024	30/06/2024
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	30/06/2024	30/06/2024
							F1901	Additional Surgery	A24	No Apparent Device Problem	30/06/2024	30/06/2024
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	30/06/2024	30/06/2024
							F1901	Additional Surgery	A24	No Apparent Device Problem	30/06/2024	30/06/2024
15936	2024	All other reportable incidents	47263	Cardiac resynchronization therapy implantable pacemaker	CT1685	Cardiac pacemakers	E011903	Syncope/Fainting	A070909	Oversensing		
									A071204	No Pacing		
							F23	Unexpected Medical Intervention	A070909	Oversensing		
									A071204	No Pacing		
					CT257	Implantable cardiac pacemakers	E011903	Syncope/Fainting	A070909	Oversensing		
									A071204	No Pacing		
							F23	Unexpected Medical Intervention	A070909	Oversensing		
									A071204	No Pacing		
					CT2866	Implantable cardiac electrical stimulators	E011903	Syncope/Fainting	A070909	Oversensing		
									A071204	No Pacing		
							F23	Unexpected Medical Intervention	A070909	Oversensing		
									A071204	No Pacing		
15937	2024	All other reportable incidents	47270	Cardiac resynchronization therapy implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A072001	Electromagnetic Interference		
									A16	Protective Measures Problem		
							F26	No Health Consequences or Impact	A072001	Electromagnetic Interference		
									A16	Protective Measures Problem		
15938	2024	All other reportable incidents	37265	Dual-chamber implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A072201	High Impedance		
							F26	No Health Consequences or Impact	A072201	High Impedance		
15939	2024	All other reportable incidents	17882	Physiologic-monitoring defibrillation system	CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A1603	Failure of Device to Self-Test		
							F27	Problem Identified During Non-Clinical Procedure	A1603	Failure of Device to Self-Test		
15940	2024	All other reportable incidents	17882	Physiologic-monitoring defibrillation system	CT1684	Cardiac pacemakers and associated devices	E0623	Unspecified Heart Problem	A0907	No Device Output		
							F05	Delay to Treatment/ Therapy	A0907	No Device Output		

15941	2024	All other reportable incidents	47267	Single-chamber implantable pacemaker, rate-responsive	CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A0908	Incorrect, Inadequate or Imprecise Result or Readings	09/09/2024	09/09/2024
							F08	Hospitalization or Prolonged Hospitalization	A0908	Incorrect, Inadequate or Imprecise Result or Readings	09/09/2024	09/09/2024
							F1905	Device Revision or Replacement	A0908	Incorrect, Inadequate or Imprecise Result or Readings	09/09/2024	09/09/2024
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A0908	Incorrect, Inadequate or Imprecise Result or Readings	09/09/2024	09/09/2024
							F08	Hospitalization or Prolonged Hospitalization	A0908	Incorrect, Inadequate or Imprecise Result or Readings	09/09/2024	09/09/2024
							F1905	Device Revision or Replacement	A0908	Incorrect, Inadequate or Imprecise Result or Readings	09/09/2024	09/09/2024
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A0908	Incorrect, Inadequate or Imprecise Result or Readings	09/09/2024	09/09/2024
							F08	Hospitalization or Prolonged Hospitalization	A0908	Incorrect, Inadequate or Imprecise Result or Readings	09/09/2024	09/09/2024
							F1905	Device Revision or Replacement	A0908	Incorrect, Inadequate or Imprecise Result or Readings	09/09/2024	09/09/2024
15942	2024		47265	Dual-chamber implantable pacemaker, rate-responsive	CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A090801	Signal Artifact/Noise		
							F25	Unanticipated Adverse Device Effect	A090801	Signal Artifact/Noise		
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A090801	Signal Artifact/Noise		
							F25	Unanticipated Adverse Device Effect	A090801	Signal Artifact/Noise		
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A090801	Signal Artifact/Noise		
							F25	Unanticipated Adverse Device Effect	A090801	Signal Artifact/Noise		
15943	2024	All other reportable incidents	17882	Physiologic-monitoring defibrillation system	CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A1603	Failure of Device to Self-Test		
							F26	No Health Consequences or Impact	A1603	Failure of Device to Self-Test		
15944	2024	All other reportable incidents	17882	Physiologic-monitoring defibrillation system	CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A1603	Failure of Device to Self-Test		
							F26	No Health Consequences or Impact	A1603	Failure of Device to Self-Test		
15945	2024	All other reportable incidents	47270	Cardiac resynchronization therapy implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A070912	Incorrect Interpretation of Signal		
									A071302	Inappropriate/Inadequate Shock/Stimulation		
							F12	Serious Injury/ Illness/ Impairment	A070912	Incorrect Interpretation of Signal		
									A071302	Inappropriate/Inadequate Shock/Stimulation		
15946	2024	All other reportable incidents	47265	Dual-chamber implantable pacemaker, rate-responsive	CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	17/05/2024	17/05/2024
									A071102	Failure to Interrogate	17/05/2024	17/05/2024
									A0723	Unexpected Mode Switch	17/05/2024	17/05/2024
							F1901	Additional Surgery	A070504	Premature Discharge of Battery	17/05/2024	17/05/2024
									A071102	Failure to Interrogate	17/05/2024	17/05/2024
									A0723	Unexpected Mode Switch	17/05/2024	17/05/2024
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	17/05/2024	17/05/2024
									A071102	Failure to Interrogate	17/05/2024	17/05/2024
									A0723	Unexpected Mode Switch	17/05/2024	17/05/2024
							F1901	Additional Surgery	A070504	Premature Discharge of Battery	17/05/2024	17/05/2024
									A071102	Failure to Interrogate	17/05/2024	17/05/2024
									A0723	Unexpected Mode Switch	17/05/2024	17/05/2024
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	17/05/2024	17/05/2024
									A071102	Failure to Interrogate	17/05/2024	17/05/2024
									A0723	Unexpected Mode Switch	17/05/2024	17/05/2024
							F1901	Additional Surgery	A070504	Premature Discharge of Battery	17/05/2024	17/05/2024
									A071102	Failure to Interrogate	17/05/2024	17/05/2024
									A0723	Unexpected Mode Switch	17/05/2024	17/05/2024
15947	2024	All other reportable incidents	47265	Dual-chamber implantable pacemaker, rate-responsive	CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem	18/09/2024	18/09/2024
									A070504	Premature Discharge of Battery	18/09/2024	18/09/2024
							F08	Hospitalization or Prolonged Hospitalization	A0705	Battery Problem	18/09/2024	18/09/2024
									A070504	Premature Discharge of Battery	18/09/2024	18/09/2024
							F1905	Device Revision or Replacement	A0705	Battery Problem	18/09/2024	18/09/2024
									A070504	Premature Discharge of Battery	18/09/2024	18/09/2024
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem	18/09/2024	18/09/2024
									A070504	Premature Discharge of Battery	18/09/2024	18/09/2024
							F08	Hospitalization or Prolonged Hospitalization	A0705	Battery Problem	18/09/2024	18/09/2024
									A070504	Premature Discharge of Battery	18/09/2024	18/09/2024
							F1905	Device Revision or Replacement	A0705	Battery Problem	18/09/2024	18/09/2024
									A070504	Premature Discharge of Battery	18/09/2024	18/09/2024
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem	18/09/2024	18/09/2024
									A070504	Premature Discharge of Battery	18/09/2024	18/09/2024
							F08	Hospitalization or Prolonged Hospitalization	A0705	Battery Problem	18/09/2024	18/09/2024
									A070504	Premature Discharge of Battery	18/09/2024	18/09/2024
							F1905	Device Revision or Replacement	A0705	Battery Problem	18/09/2024	18/09/2024
									A070504	Premature Discharge of Battery	18/09/2024	18/09/2024
15948	2024	All other reportable incidents	47270	Cardiac resynchronization therapy implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E0618	Non Specific EKG/ECG Changes	A070909	Oversensing		
									A0712	Pacing Problem		
									A072001	Electromagnetic Interference		
									A090801	Signal Artifact/Noise		
							F11	Non-Serious Injury/ Illness/ Impairment	A070909	Oversensing		
									A0712	Pacing Problem		
									A072001	Electromagnetic Interference		
									A090801	Signal Artifact/Noise		
15949	2024	All other reportable incidents	47270	Cardiac resynchronization therapy implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	05/11/2024	05/11/2024
							F1901	Additional Surgery	A070504	Premature Discharge of Battery	05/11/2024	05/11/2024
15950	2024		47265	Dual-chamber implantable pacemaker, rate-responsive	CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	13/09/2024	13/09/2024
							F1901	Additional Surgery	A24	No Apparent Device Problem	13/09/2024	13/09/2024
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	13/09/2024	13/09/2024
							F1901	Additional Surgery	A24	No Apparent Device Problem	13/09/2024	13/09/2024
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	13/09/2024	13/09/2024
							F1901	Additional Surgery	A24	No Apparent Device Problem	13/09/2024	13/09/2024
15951	2024	All other reportable incidents	47265	Dual-chamber implantable pacemaker, rate-responsive	CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A070101	Failure to Capture		
									A0709	Device Sensing Problem		
							F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A070101	Failure to Capture		
									A0709	Device Sensing Problem		
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A070101	Failure to Capture		
									A0709	Device Sensing Problem		

							F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A070101	Failure to Capture		
									A0709	Device Sensing Problem		
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A070101	Failure to Capture		
									A0709	Device Sensing Problem		
							F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A070101	Failure to Capture		
									A0709	Device Sensing Problem		
15952	2024	All other reportable incidents	47270	Cardiac resynchronization therapy implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E210401	Shock from Patient Lead(s)	A040101	Fracture	10/10/2024	10/10/2024
									A070909	Oversensing	10/10/2024	10/10/2024
									A071302	Inappropriate/Inadequate Shock/Stimulation	10/10/2024	10/10/2024
									A090801	Signal Artifact/Noise	10/10/2024	10/10/2024
									A12	Connection Problem	10/10/2024	10/10/2024
									A1301	Failure to Read Input Signal	10/10/2024	10/10/2024
							F08	Hospitalization or Prolonged Hospitalization	A040101	Fracture	10/10/2024	10/10/2024
									A070909	Oversensing	10/10/2024	10/10/2024
									A071302	Inappropriate/Inadequate Shock/Stimulation	10/10/2024	10/10/2024
									A090801	Signal Artifact/Noise	10/10/2024	10/10/2024
									A12	Connection Problem	10/10/2024	10/10/2024
									A1301	Failure to Read Input Signal	10/10/2024	10/10/2024
							F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A040101	Fracture	10/10/2024	10/10/2024
									A070909	Oversensing	10/10/2024	10/10/2024
									A071302	Inappropriate/Inadequate Shock/Stimulation	10/10/2024	10/10/2024
									A090801	Signal Artifact/Noise	10/10/2024	10/10/2024
									A12	Connection Problem	10/10/2024	10/10/2024
									A1301	Failure to Read Input Signal	10/10/2024	10/10/2024
							F1905	Device Revision or Replacement	A040101	Fracture	10/10/2024	10/10/2024
									A070909	Oversensing	10/10/2024	10/10/2024
									A071302	Inappropriate/Inadequate Shock/Stimulation	10/10/2024	10/10/2024
									A090801	Signal Artifact/Noise	10/10/2024	10/10/2024
									A12	Connection Problem	10/10/2024	10/10/2024
									A1301	Failure to Read Input Signal	10/10/2024	10/10/2024
							F2203	Imaging Required	A040101	Fracture	10/10/2024	10/10/2024
									A070909	Oversensing	10/10/2024	10/10/2024
									A071302	Inappropriate/Inadequate Shock/Stimulation	10/10/2024	10/10/2024
									A090801	Signal Artifact/Noise	10/10/2024	10/10/2024
									A12	Connection Problem	10/10/2024	10/10/2024
									A1301	Failure to Read Input Signal	10/10/2024	10/10/2024
15953	2024	All other reportable incidents	17882	Physiologic-monitoring defibrillation system	CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A0708	Power Problem		
							F27	Problem Identified During Non-Clinical Procedure	A0708	Power Problem		
15954	2024	All other reportable incidents	17882	Physiologic-monitoring defibrillation system	CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A0908	Incorrect, Inadequate or Imprecise Result or Readings		
							F27	Problem Identified During Non-Clinical Procedure	A0908	Incorrect, Inadequate or Imprecise Result or Readings		
15955	2024	All other reportable incidents	35852	Single-chamber implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A070909	Oversensing		
							F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A070909	Oversensing		
15956	2024	All other reportable incidents	47265	Dual-chamber implantable pacemaker, rate-responsive	CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	25/09/2024	25/09/2024
									A0723	Unexpected Mode Switch	25/09/2024	25/09/2024
							F1901	Additional Surgery	A070504	Premature Discharge of Battery	25/09/2024	25/09/2024
									A0723	Unexpected Mode Switch	25/09/2024	25/09/2024
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	25/09/2024	25/09/2024
									A0723	Unexpected Mode Switch	25/09/2024	25/09/2024
							F1901	Additional Surgery	A070504	Premature Discharge of Battery	25/09/2024	25/09/2024
									A0723	Unexpected Mode Switch	25/09/2024	25/09/2024
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	25/09/2024	25/09/2024
									A0723	Unexpected Mode Switch	25/09/2024	25/09/2024
							F1901	Additional Surgery	A070504	Premature Discharge of Battery	25/09/2024	25/09/2024
									A0723	Unexpected Mode Switch	25/09/2024	25/09/2024
15957	2024	All other reportable incidents	47263	Cardiac resynchronization therapy implantable pacemaker	CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A070909	Oversensing		
							F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A070909	Oversensing		
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A070909	Oversensing		
							F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A070909	Oversensing		
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A070909	Oversensing		
							F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A070909	Oversensing		
15958	2024	All other reportable incidents	47263	Cardiac resynchronization therapy implantable pacemaker	CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A0203	Defective Device	13/07/2023	13/07/2023
							F24	Insufficient Health Impact Information	A0203	Defective Device	13/07/2023	13/07/2023
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A0203	Defective Device	13/07/2023	13/07/2023
							F24	Insufficient Health Impact Information	A0203	Defective Device	13/07/2023	13/07/2023
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A0203	Defective Device	13/07/2023	13/07/2023
							F24	Insufficient Health Impact Information	A0203	Defective Device	13/07/2023	13/07/2023
15959	2024		47267	Single-chamber implantable pacemaker, rate-responsive	CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A071102	Failure to Interrogate	05/09/2024	05/09/2024
							F1901	Additional Surgery	A071102	Failure to Interrogate	05/09/2024	05/09/2024
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A071102	Failure to Interrogate	05/09/2024	05/09/2024
							F1901	Additional Surgery	A071102	Failure to Interrogate	05/09/2024	05/09/2024
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A071102	Failure to Interrogate	05/09/2024	05/09/2024
							F1901	Additional Surgery	A071102	Failure to Interrogate	05/09/2024	05/09/2024
15960	2024	All other reportable incidents	60789	Intracardiac pacemaker	CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A040601	Deformation Due to Compressive Stress	14/09/2024	14/09/2024
									A070101	Failure to Capture	14/09/2024	14/09/2024
									A070102	High Capture Threshold	14/09/2024	14/09/2024
									A0712	Pacing Problem	14/09/2024	14/09/2024
									A1502	Positioning Problem	14/09/2024	14/09/2024
									A23	Use of Device Problem	14/09/2024	14/09/2024
							F26	No Health Consequences or Impact	A040601	Deformation Due to Compressive Stress	14/09/2024	14/09/2024
									A070101	Failure to Capture	14/09/2024	14/09/2024
									A070102	High Capture Threshold	14/09/2024	14/09/2024
									A0712	Pacing Problem	14/09/2024	14/09/2024
									A1502	Positioning Problem	14/09/2024	14/09/2024
									A23	Use of Device Problem	14/09/2024	14/09/2024
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A040601	Deformation Due to Compressive Stress	14/09/2024	14/09/2024
									A070101	Failure to Capture	14/09/2024	14/09/2024

									A070102	High Capture Threshold	14/09/2024	14/09/2024
									A0712	Pacing Problem	14/09/2024	14/09/2024
									A1502	Positioning Problem	14/09/2024	14/09/2024
									A23	Use of Device Problem	14/09/2024	14/09/2024
							F26	No Health Consequences or Impact	A040601	Deformation Due to Compressive Stress	14/09/2024	14/09/2024
									A070101	Failure to Capture	14/09/2024	14/09/2024
									A070102	High Capture Threshold	14/09/2024	14/09/2024
									A0712	Pacing Problem	14/09/2024	14/09/2024
									A1502	Positioning Problem	14/09/2024	14/09/2024
									A23	Use of Device Problem	14/09/2024	14/09/2024
					CT2818	Catheterization/Cannulation/Fluid path devices	E2403	No Clinical Signs, Symptoms or Conditions	A040601	Deformation Due to Compressive Stress	14/09/2024	14/09/2024
									A070101	Failure to Capture	14/09/2024	14/09/2024
									A070102	High Capture Threshold	14/09/2024	14/09/2024
									A0712	Pacing Problem	14/09/2024	14/09/2024
									A1502	Positioning Problem	14/09/2024	14/09/2024
									A23	Use of Device Problem	14/09/2024	14/09/2024
							F26	No Health Consequences or Impact	A040601	Deformation Due to Compressive Stress	14/09/2024	14/09/2024
									A070101	Failure to Capture	14/09/2024	14/09/2024
									A070102	High Capture Threshold	14/09/2024	14/09/2024
									A0712	Pacing Problem	14/09/2024	14/09/2024
									A1502	Positioning Problem	14/09/2024	14/09/2024
									A23	Use of Device Problem	14/09/2024	14/09/2024
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A040601	Deformation Due to Compressive Stress	14/09/2024	14/09/2024
									A070101	Failure to Capture	14/09/2024	14/09/2024
									A070102	High Capture Threshold	14/09/2024	14/09/2024
									A0712	Pacing Problem	14/09/2024	14/09/2024
									A1502	Positioning Problem	14/09/2024	14/09/2024
									A23	Use of Device Problem	14/09/2024	14/09/2024
							F26	No Health Consequences or Impact	A040601	Deformation Due to Compressive Stress	14/09/2024	14/09/2024
									A070101	Failure to Capture	14/09/2024	14/09/2024
									A070102	High Capture Threshold	14/09/2024	14/09/2024
									A0712	Pacing Problem	14/09/2024	14/09/2024
									A1502	Positioning Problem	14/09/2024	14/09/2024
									A23	Use of Device Problem	14/09/2024	14/09/2024
15961	2024	All other reportable incidents	17882	Physiologic-monitoring defibrillation system	CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A0719	Unexpected Shutdown		
							F27	Problem Identified During Non-Clinical Procedure	A0719	Unexpected Shutdown		
15962	2024	All other reportable incidents	47205	Cardiac pulse generator programmer	CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A11	Computer Software Problem		
									A13	Communication or Transmission Problem		
							F26	No Health Consequences or Impact	A11	Computer Software Problem		
									A13	Communication or Transmission Problem		
15963	2024	All other reportable incidents	47265	Dual-chamber implantable pacemaker, rate-responsive	CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A070909	Oversensing		
									A072201	High Impedance		
									A090801	Signal Artifact/Noise		
							F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A070909	Oversensing		
									A072201	High Impedance		
									A090801	Signal Artifact/Noise		
							F2203	Imaging Required	A070909	Oversensing		
									A072201	High Impedance		
									A090801	Signal Artifact/Noise		
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A070909	Oversensing		
									A072201	High Impedance		
									A090801	Signal Artifact/Noise		
							F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A070909	Oversensing		
									A072201	High Impedance		
									A090801	Signal Artifact/Noise		
							F2203	Imaging Required	A070909	Oversensing		
									A072201	High Impedance		
									A090801	Signal Artifact/Noise		
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A070909	Oversensing		
									A072201	High Impedance		
									A090801	Signal Artifact/Noise		
							F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A070909	Oversensing		
									A072201	High Impedance		
									A090801	Signal Artifact/Noise		
							F2203	Imaging Required	A070909	Oversensing		
									A072201	High Impedance		
									A090801	Signal Artifact/Noise		
15964	2024	All other reportable incidents	35852	Single-chamber implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A070909	Oversensing		
									A072001	Electromagnetic Interference		
									A090801	Signal Artifact/Noise		
							F12	Serious Injury/ Illness/ Impairment	A070909	Oversensing		
									A072001	Electromagnetic Interference		
									A090801	Signal Artifact/Noise		
							F2307	Reprogramming of Device	A070909	Oversensing		
									A072001	Electromagnetic Interference		
									A090801	Signal Artifact/Noise		
15965	2024	All other reportable incidents	35852	Single-chamber implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A072201	High Impedance		
							F26	No Health Consequences or Impact	A072201	High Impedance		
15966	2024	All other reportable incidents	37265	Dual-chamber implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A0712	Pacing Problem	21/08/2024	21/08/2024
									A16	Protective Measures Problem	21/08/2024	21/08/2024
							F08	Hospitalization or Prolonged Hospitalization	A0712	Pacing Problem	21/08/2024	21/08/2024
									A16	Protective Measures Problem	21/08/2024	21/08/2024
							F12	Serious Injury/ Illness/ Impairment	A0712	Pacing Problem	21/08/2024	21/08/2024
									A16	Protective Measures Problem	21/08/2024	21/08/2024
							F1905	Device Revision or Replacement	A0712	Pacing Problem	21/08/2024	21/08/2024
									A16	Protective Measures Problem	21/08/2024	21/08/2024
15967	2024	All other reportable incidents	47205	Cardiac pulse generator programmer	CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A13	Communication or Transmission Problem		
							F27	Problem Identified During Non-Clinical Procedure	A13	Communication or Transmission Problem		

15968	2024	All other reportable incidents	47205	Cardiac pulse generator programmer	CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A0902	Display or Visual Feedback Problem		
							F27	Problem Identified During Non-Clinical Procedure	A0902	Display or Visual Feedback Problem		
15969	2024	All other reportable incidents	47205	Cardiac pulse generator programmer	CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A13	Communication or Transmission Problem		
							F23	Unexpected Medical Intervention	A13	Communication or Transmission Problem		
15970	2024	All other reportable incidents	47265	Dual-chamber implantable pacemaker, rate-responsive	CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F1901	Additional Surgery	A070504	Premature Discharge of Battery		
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F1901	Additional Surgery	A070504	Premature Discharge of Battery		
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F1901	Additional Surgery	A070504	Premature Discharge of Battery		
15971	2024	All other reportable incidents	47265	Dual-chamber implantable pacemaker, rate-responsive	CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	18/09/2024	18/09/2024
							F1901	Additional Surgery	A070504	Premature Discharge of Battery	18/09/2024	18/09/2024
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	18/09/2024	18/09/2024
							F1901	Additional Surgery	A070504	Premature Discharge of Battery	18/09/2024	18/09/2024
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	18/09/2024	18/09/2024
							F1901	Additional Surgery	A070504	Premature Discharge of Battery	18/09/2024	18/09/2024
15972	2024	All other reportable incidents	47265	Dual-chamber implantable pacemaker, rate-responsive	CT1685	Cardiac pacemakers	E0112	Dizziness	A0705	Battery Problem		
							F26	No Health Consequences or Impact	A0705	Battery Problem		
					CT257	Implantable cardiac pacemakers	E0112	Dizziness	A0705	Battery Problem		
							F26	No Health Consequences or Impact	A0705	Battery Problem		
					CT2866	Implantable cardiac electrical stimulators	E0112	Dizziness	A0705	Battery Problem		
							F26	No Health Consequences or Impact	A0705	Battery Problem		
15973	2024	All other reportable incidents	47265	Dual-chamber implantable pacemaker, rate-responsive	CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem		
							F1901	Additional Surgery	A24	No Apparent Device Problem		
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem		
							F1901	Additional Surgery	A24	No Apparent Device Problem		
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem		
							F1901	Additional Surgery	A24	No Apparent Device Problem		
15974	2024	All other reportable incidents	17882	Physiologic-monitoring defibrillation system	CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A0719	Unexpected Shutdown		
							F27	Problem Identified During Non-Clinical Procedure	A0719	Unexpected Shutdown		
15975	2024	All other reportable incidents	47270	Cardiac resynchronization therapy implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A070909	Oversensing		
							F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A070909	Oversensing		
15976	2024	All other reportable incidents	17882	Physiologic-monitoring defibrillation system	CT1684	Cardiac pacemakers and associated devices	E0623	Unspecified Heart Problem	A071301	Failure to Deliver Shock/Stimulation		
							F05	Delay to Treatment/ Therapy	A071301	Failure to Deliver Shock/Stimulation		
15977	2024	All other reportable incidents	17882	Physiologic-monitoring defibrillation system	CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A1603	Failure of Device to Self-Test		
							F26	No Health Consequences or Impact	A1603	Failure of Device to Self-Test		
15978	2024	All other reportable incidents	17882	Physiologic-monitoring defibrillation system	CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A1603	Failure of Device to Self-Test		
							F27	Problem Identified During Non-Clinical Procedure	A1603	Failure of Device to Self-Test		
15979	2024	All other reportable incidents	17882	Physiologic-monitoring defibrillation system	CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A071301	Failure to Deliver Shock/Stimulation		
									A0719	Unexpected Shutdown		
							F26	No Health Consequences or Impact	A071301	Failure to Deliver Shock/Stimulation		
									A0719	Unexpected Shutdown		
15980	2024		47265	Dual-chamber implantable pacemaker, rate-responsive	CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	24/04/2024	24/04/2024
							F1901	Additional Surgery	A24	No Apparent Device Problem	24/04/2024	24/04/2024
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	24/04/2024	24/04/2024
							F1901	Additional Surgery	A24	No Apparent Device Problem	24/04/2024	24/04/2024
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	24/04/2024	24/04/2024
							F1901	Additional Surgery	A24	No Apparent Device Problem	24/04/2024	24/04/2024
15981	2024	All other reportable incidents	37265	Dual-chamber implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A0709	Device Sensing Problem		
									A070909	Oversensing		
									A072001	Electromagnetic Interference		
									A090801	Signal Artifact/Noise		
									A160105	False Alarm		
							F12	Serious Injury/ Illness/ Impairment	A0709	Device Sensing Problem		
									A070909	Oversensing		
									A072001	Electromagnetic Interference		
									A090801	Signal Artifact/Noise		
									A160105	False Alarm		
							F2307	Reprogramming of Device	A0709	Device Sensing Problem		
									A070909	Oversensing		
									A072001	Electromagnetic Interference		
									A090801	Signal Artifact/Noise		
									A160105	False Alarm		
15982	2024	All other reportable incidents	37265	Dual-chamber implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A05	Mechanical Problem	26/09/2024	26/09/2024
							F26	No Health Consequences or Impact	A05	Mechanical Problem	26/09/2024	26/09/2024
15983	2024	All other reportable incidents	47265	Dual-chamber implantable pacemaker, rate-responsive	CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem	20/09/2024	20/09/2024
									A070504	Premature Discharge of Battery	20/09/2024	20/09/2024
							F08	Hospitalization or Prolonged Hospitalization	A0705	Battery Problem	20/09/2024	20/09/2024
									A070504	Premature Discharge of Battery	20/09/2024	20/09/2024
							F1905	Device Revision or Replacement	A0705	Battery Problem	20/09/2024	20/09/2024
									A070504	Premature Discharge of Battery	20/09/2024	20/09/2024
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem	20/09/2024	20/09/2024
									A070504	Premature Discharge of Battery	20/09/2024	20/09/2024
							F08	Hospitalization or Prolonged Hospitalization	A0705	Battery Problem	20/09/2024	20/09/2024
									A070504	Premature Discharge of Battery	20/09/2024	20/09/2024
							F1905	Device Revision or Replacement	A0705	Battery Problem	20/09/2024	20/09/2024
									A070504	Premature Discharge of Battery	20/09/2024	20/09/2024
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem	20/09/2024	20/09/2024
									A070504	Premature Discharge of Battery	20/09/2024	20/09/2024
							F08	Hospitalization or Prolonged Hospitalization	A0705	Battery Problem	20/09/2024	20/09/2024
									A070504	Premature Discharge of Battery	20/09/2024	20/09/2024
							F1905	Device Revision or Replacement	A0705	Battery Problem	20/09/2024	20/09/2024
									A070504	Premature Discharge of Battery	20/09/2024	20/09/2024
15984	2024	All other reportable incidents	47263	Cardiac resynchronization therapy implantable pacemaker	CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A070909	Oversensing		
									A090801	Signal Artifact/Noise		
									A1301	Failure to Read Input Signal		
							F2203	Imaging Required	A070909	Oversensing		
									A090801	Signal Artifact/Noise		

									A1301	Failure to Read Input Signal		
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A070909	Oversensing		
									A090801	Signal Artifact/Noise		
									A1301	Failure to Read Input Signal		
							F2203	Imaging Required	A070909	Oversensing		
									A090801	Signal Artifact/Noise		
									A1301	Failure to Read Input Signal		
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A070909	Oversensing		
									A090801	Signal Artifact/Noise		
									A1301	Failure to Read Input Signal		
							F2203	Imaging Required	A070909	Oversensing		
									A090801	Signal Artifact/Noise		
									A1301	Failure to Read Input Signal		
15985	2024	All other reportable incidents			CT1684	Cardiac pacemakers and associated devices	E060104	Bradycardia	A070909	Oversensing		
									A0712	Pacing Problem		
									A072201	High Impedance		
							F1001	Absence of Treatment	A070909	Oversensing		
									A0712	Pacing Problem		
									A072201	High Impedance		
15986	2024	All other reportable incidents	47205	Cardiac pulse generator programmer	CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A150105	Key or Button Unresponsive/not Working		
							F27	Problem Identified During Non-Clinical Procedure	A150105	Key or Button Unresponsive/not Working		
15987	2024	All other reportable incidents	47265	Dual-chamber implantable pacemaker, rate-responsive	CT1685	Cardiac pacemakers	E0402	Hypersensitivity/Allergic Reaction	A24	No Apparent Device Problem	26/04/2024	26/04/2024
							E1906	Unspecified Infection	A24	No Apparent Device Problem	26/04/2024	26/04/2024
							E2006	Erosion	A24	No Apparent Device Problem	26/04/2024	26/04/2024
							F08	Hospitalization or Prolonged Hospitalization	A24	No Apparent Device Problem	26/04/2024	26/04/2024
							F12	Serious Injury/ Illness/ Impairment	A24	No Apparent Device Problem	26/04/2024	26/04/2024
							F1903	Device Explantation	A24	No Apparent Device Problem	26/04/2024	26/04/2024
					CT257	Implantable cardiac pacemakers	E0402	Hypersensitivity/Allergic Reaction	A24	No Apparent Device Problem	26/04/2024	26/04/2024
							E1906	Unspecified Infection	A24	No Apparent Device Problem	26/04/2024	26/04/2024
							E2006	Erosion	A24	No Apparent Device Problem	26/04/2024	26/04/2024
							F08	Hospitalization or Prolonged Hospitalization	A24	No Apparent Device Problem	26/04/2024	26/04/2024
							F12	Serious Injury/ Illness/ Impairment	A24	No Apparent Device Problem	26/04/2024	26/04/2024
							F1903	Device Explantation	A24	No Apparent Device Problem	26/04/2024	26/04/2024
					CT2866	Implantable cardiac electrical stimulators	E0402	Hypersensitivity/Allergic Reaction	A24	No Apparent Device Problem	26/04/2024	26/04/2024
							E1906	Unspecified Infection	A24	No Apparent Device Problem	26/04/2024	26/04/2024
							E2006	Erosion	A24	No Apparent Device Problem	26/04/2024	26/04/2024
							F08	Hospitalization or Prolonged Hospitalization	A24	No Apparent Device Problem	26/04/2024	26/04/2024
							F12	Serious Injury/ Illness/ Impairment	A24	No Apparent Device Problem	26/04/2024	26/04/2024
							F1903	Device Explantation	A24	No Apparent Device Problem	26/04/2024	26/04/2024
15988	2024	All other reportable incidents	47265	Dual-chamber implantable pacemaker, rate-responsive	CT1685	Cardiac pacemakers	E0402	Hypersensitivity/Allergic Reaction	A24	No Apparent Device Problem	26/04/2024	26/04/2024
							E1906	Unspecified Infection	A24	No Apparent Device Problem	26/04/2024	26/04/2024
							E2006	Erosion	A24	No Apparent Device Problem	26/04/2024	26/04/2024
							F08	Hospitalization or Prolonged Hospitalization	A24	No Apparent Device Problem	26/04/2024	26/04/2024
							F12	Serious Injury/ Illness/ Impairment	A24	No Apparent Device Problem	26/04/2024	26/04/2024
							F1903	Device Explantation	A24	No Apparent Device Problem	26/04/2024	26/04/2024
					CT257	Implantable cardiac pacemakers	E0402	Hypersensitivity/Allergic Reaction	A24	No Apparent Device Problem	26/04/2024	26/04/2024
							E1906	Unspecified Infection	A24	No Apparent Device Problem	26/04/2024	26/04/2024
							E2006	Erosion	A24	No Apparent Device Problem	26/04/2024	26/04/2024
							F08	Hospitalization or Prolonged Hospitalization	A24	No Apparent Device Problem	26/04/2024	26/04/2024
							F12	Serious Injury/ Illness/ Impairment	A24	No Apparent Device Problem	26/04/2024	26/04/2024
							F1903	Device Explantation	A24	No Apparent Device Problem	26/04/2024	26/04/2024
					CT2866	Implantable cardiac electrical stimulators	E0402	Hypersensitivity/Allergic Reaction	A24	No Apparent Device Problem	26/04/2024	26/04/2024
							E1906	Unspecified Infection	A24	No Apparent Device Problem	26/04/2024	26/04/2024
							E2006	Erosion	A24	No Apparent Device Problem	26/04/2024	26/04/2024
							F08	Hospitalization or Prolonged Hospitalization	A24	No Apparent Device Problem	26/04/2024	26/04/2024
							F12	Serious Injury/ Illness/ Impairment	A24	No Apparent Device Problem	26/04/2024	26/04/2024
							F1903	Device Explantation	A24	No Apparent Device Problem	26/04/2024	26/04/2024
15989	2024	All other reportable incidents	60789	Intracardiac pacemaker	CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A070102	High Capture Threshold	27/09/2024	27/09/2024
									A1502	Positioning Problem	27/09/2024	27/09/2024
									A23	Use of Device Problem	27/09/2024	27/09/2024
							F26	No Health Consequences or Impact	A070102	High Capture Threshold	27/09/2024	27/09/2024
									A1502	Positioning Problem	27/09/2024	27/09/2024
									A23	Use of Device Problem	27/09/2024	27/09/2024
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A070102	High Capture Threshold	27/09/2024	27/09/2024
									A1502	Positioning Problem	27/09/2024	27/09/2024
									A23	Use of Device Problem	27/09/2024	27/09/2024
							F26	No Health Consequences or Impact	A070102	High Capture Threshold	27/09/2024	27/09/2024
									A1502	Positioning Problem	27/09/2024	27/09/2024
									A23	Use of Device Problem	27/09/2024	27/09/2024
					CT2818	Catheterization/Cannulation/Fluid path devices	E2403	No Clinical Signs, Symptoms or Conditions	A070102	High Capture Threshold	27/09/2024	27/09/2024
									A1502	Positioning Problem	27/09/2024	27/09/2024
									A23	Use of Device Problem	27/09/2024	27/09/2024
							F26	No Health Consequences or Impact	A070102	High Capture Threshold	27/09/2024	27/09/2024
									A1502	Positioning Problem	27/09/2024	27/09/2024
									A23	Use of Device Problem	27/09/2024	27/09/2024
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A070102	High Capture Threshold	27/09/2024	27/09/2024
									A1502	Positioning Problem	27/09/2024	27/09/2024
									A23	Use of Device Problem	27/09/2024	27/09/2024
							F26	No Health Consequences or Impact	A070102	High Capture Threshold	27/09/2024	27/09/2024
									A1502	Positioning Problem	27/09/2024	27/09/2024
									A23	Use of Device Problem	27/09/2024	27/09/2024
15990	2024	All other reportable incidents	60789	Intracardiac pacemaker	CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A070102	High Capture Threshold	27/09/2024	27/09/2024
									A1502	Positioning Problem	27/09/2024	27/09/2024
									A23	Use of Device Problem	27/09/2024	27/09/2024
							F26	No Health Consequences or Impact	A070102	High Capture Threshold	27/09/2024	27/09/2024
									A1502	Positioning Problem	27/09/2024	27/09/2024
									A23	Use of Device Problem	27/09/2024	27/09/2024
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A070102	High Capture Threshold	27/09/2024	27/09/2024

									A1502	Positioning Problem	27/09/2024	27/09/2024
									A23	Use of Device Problem	27/09/2024	27/09/2024
							F26	No Health Consequences or Impact	A070102	High Capture Threshold	27/09/2024	27/09/2024
									A1502	Positioning Problem	27/09/2024	27/09/2024
									A23	Use of Device Problem	27/09/2024	27/09/2024
					CT2818	Catheterization/Cannulation/Fluid path devices	E2403	No Clinical Signs, Symptoms or Conditions	A070102	High Capture Threshold	27/09/2024	27/09/2024
									A1502	Positioning Problem	27/09/2024	27/09/2024
									A23	Use of Device Problem	27/09/2024	27/09/2024
							F26	No Health Consequences or Impact	A070102	High Capture Threshold	27/09/2024	27/09/2024
									A1502	Positioning Problem	27/09/2024	27/09/2024
									A23	Use of Device Problem	27/09/2024	27/09/2024
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A070102	High Capture Threshold	27/09/2024	27/09/2024
									A1502	Positioning Problem	27/09/2024	27/09/2024
									A23	Use of Device Problem	27/09/2024	27/09/2024
							F26	No Health Consequences or Impact	A070102	High Capture Threshold	27/09/2024	27/09/2024
									A1502	Positioning Problem	27/09/2024	27/09/2024
									A23	Use of Device Problem	27/09/2024	27/09/2024
15991	2024	All other reportable incidents	60789	Intracardiac pacemaker	CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A040601	Deformation Due to Compressive Stress	27/09/2024	27/09/2024
									A07	Electrical /Electronic Property Problem	27/09/2024	27/09/2024
									A070101	Failure to Capture	27/09/2024	27/09/2024
									A070102	High Capture Threshold	27/09/2024	27/09/2024
									A070901	Decreased Sensitivity	27/09/2024	27/09/2024
							F26	No Health Consequences or Impact	A040601	Deformation Due to Compressive Stress	27/09/2024	27/09/2024
									A07	Electrical /Electronic Property Problem	27/09/2024	27/09/2024
									A070101	Failure to Capture	27/09/2024	27/09/2024
									A070102	High Capture Threshold	27/09/2024	27/09/2024
									A070901	Decreased Sensitivity	27/09/2024	27/09/2024
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A040601	Deformation Due to Compressive Stress	27/09/2024	27/09/2024
									A07	Electrical /Electronic Property Problem	27/09/2024	27/09/2024
									A070101	Failure to Capture	27/09/2024	27/09/2024
									A070102	High Capture Threshold	27/09/2024	27/09/2024
									A070901	Decreased Sensitivity	27/09/2024	27/09/2024
							F26	No Health Consequences or Impact	A040601	Deformation Due to Compressive Stress	27/09/2024	27/09/2024
									A07	Electrical /Electronic Property Problem	27/09/2024	27/09/2024
									A070101	Failure to Capture	27/09/2024	27/09/2024
									A070102	High Capture Threshold	27/09/2024	27/09/2024
									A070901	Decreased Sensitivity	27/09/2024	27/09/2024
					CT2818	Catheterization/Cannulation/Fluid path devices	E2403	No Clinical Signs, Symptoms or Conditions	A040601	Deformation Due to Compressive Stress	27/09/2024	27/09/2024
									A07	Electrical /Electronic Property Problem	27/09/2024	27/09/2024
									A070101	Failure to Capture	27/09/2024	27/09/2024
									A070102	High Capture Threshold	27/09/2024	27/09/2024
									A070901	Decreased Sensitivity	27/09/2024	27/09/2024
							F26	No Health Consequences or Impact	A040601	Deformation Due to Compressive Stress	27/09/2024	27/09/2024
									A07	Electrical /Electronic Property Problem	27/09/2024	27/09/2024
									A070101	Failure to Capture	27/09/2024	27/09/2024
									A070102	High Capture Threshold	27/09/2024	27/09/2024
									A070901	Decreased Sensitivity	27/09/2024	27/09/2024
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A040601	Deformation Due to Compressive Stress	27/09/2024	27/09/2024
									A07	Electrical /Electronic Property Problem	27/09/2024	27/09/2024
									A070101	Failure to Capture	27/09/2024	27/09/2024
									A070102	High Capture Threshold	27/09/2024	27/09/2024
									A070901	Decreased Sensitivity	27/09/2024	27/09/2024
							F26	No Health Consequences or Impact	A040601	Deformation Due to Compressive Stress	27/09/2024	27/09/2024
									A07	Electrical /Electronic Property Problem	27/09/2024	27/09/2024
									A070101	Failure to Capture	27/09/2024	27/09/2024
									A070102	High Capture Threshold	27/09/2024	27/09/2024
									A070901	Decreased Sensitivity	27/09/2024	27/09/2024
15992	2024	All other reportable incidents	47270	Cardiac resynchronization therapy implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A0709	Device Sensing Problem	24/09/2024	24/09/2024
									A1206	Misconnection	24/09/2024	24/09/2024
							F26	No Health Consequences or Impact	A0709	Device Sensing Problem	24/09/2024	24/09/2024
									A1206	Misconnection	24/09/2024	24/09/2024
15993	2024	All other reportable incidents	47265	Dual-chamber implantable pacemaker, rate-responsive	CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem	25/09/2024	25/09/2024
									A070504	Premature Discharge of Battery	25/09/2024	25/09/2024
							F08	Hospitalization or Prolonged Hospitalization	A0705	Battery Problem	25/09/2024	25/09/2024
									A070504	Premature Discharge of Battery	25/09/2024	25/09/2024
							F1905	Device Revision or Replacement	A0705	Battery Problem	25/09/2024	25/09/2024
									A070504	Premature Discharge of Battery	25/09/2024	25/09/2024
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem	25/09/2024	25/09/2024
									A070504	Premature Discharge of Battery	25/09/2024	25/09/2024
							F08	Hospitalization or Prolonged Hospitalization	A0705	Battery Problem	25/09/2024	25/09/2024
									A070504	Premature Discharge of Battery	25/09/2024	25/09/2024
							F1905	Device Revision or Replacement	A0705	Battery Problem	25/09/2024	25/09/2024
									A070504	Premature Discharge of Battery	25/09/2024	25/09/2024
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem	25/09/2024	25/09/2024
									A070504	Premature Discharge of Battery	25/09/2024	25/09/2024
							F08	Hospitalization or Prolonged Hospitalization	A0705	Battery Problem	25/09/2024	25/09/2024
									A070504	Premature Discharge of Battery	25/09/2024	25/09/2024
							F1905	Device Revision or Replacement	A0705	Battery Problem	25/09/2024	25/09/2024
									A070504	Premature Discharge of Battery	25/09/2024	25/09/2024
15994	2024	All other reportable incidents	47265	Dual-chamber implantable pacemaker, rate-responsive	CT1685	Cardiac pacemakers	E020201	Anxiety	A0203	Defective Device	17/01/2024	17/01/2024
							F24	Insufficient Health Impact Information	A0203	Defective Device	17/01/2024	17/01/2024
					CT257	Implantable cardiac pacemakers	E020201	Anxiety	A0203	Defective Device	17/01/2024	17/01/2024
							F24	Insufficient Health Impact Information	A0203	Defective Device	17/01/2024	17/01/2024
					CT2866	Implantable cardiac electrical stimulators	E020201	Anxiety	A0203	Defective Device	17/01/2024	17/01/2024
							F24	Insufficient Health Impact Information	A0203	Defective Device	17/01/2024	17/01/2024
15995	2024	All other reportable incidents	47265	Dual-chamber implantable pacemaker, rate-responsive	CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
									A0908	Incorrect, Inadequate or Imprecise Result or Readings		

							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
									A0908	Incorrect, Inadequate or Imprecise Result or Readings		
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
									A0908	Incorrect, Inadequate or Imprecise Result or Readings		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
									A0908	Incorrect, Inadequate or Imprecise Result or Readings		
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
									A0908	Incorrect, Inadequate or Imprecise Result or Readings		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
									A0908	Incorrect, Inadequate or Imprecise Result or Readings		
15996	2024	All other reportable incidents	37265	Dual-chamber implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E210401	Shock from Patient Lead(s)	A0709	Device Sensing Problem		
									A071302	Inappropriate/Inadequate Shock/Stimulation		
							F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A0709	Device Sensing Problem		
									A071302	Inappropriate/Inadequate Shock/Stimulation		
							F1203	Life Threatening Illness or Injury	A0709	Device Sensing Problem		
									A071302	Inappropriate/Inadequate Shock/Stimulation		
15997	2024	All other reportable incidents	47265	Dual-chamber implantable pacemaker, rate-responsive	CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A071102	Failure to Interrogate	02/08/2024	02/08/2024
							F1901	Additional Surgery	A071102	Failure to Interrogate	02/08/2024	02/08/2024
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A071102	Failure to Interrogate	02/08/2024	02/08/2024
							F1901	Additional Surgery	A071102	Failure to Interrogate	02/08/2024	02/08/2024
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A071102	Failure to Interrogate	02/08/2024	02/08/2024
							F1901	Additional Surgery	A071102	Failure to Interrogate	02/08/2024	02/08/2024
15998	2024	All other reportable incidents	37265	Dual-chamber implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A090801	Signal Artifact/Noise		
							F1901	Additional Surgery	A090801	Signal Artifact/Noise		
15999	2024	All other reportable incidents	17882	Physiologic-monitoring defibrillation system	CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A07	Electrical /Electronic Property Problem		
							F27	Problem Identified During Non-Clinical Procedure	A07	Electrical /Electronic Property Problem		
16000	2024	All other reportable incidents	37265	Dual-chamber implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A072001	Electromagnetic Interference		
							F26	No Health Consequences or Impact	A072001	Electromagnetic Interference		
16001	2024	All other reportable incidents	35852	Single-chamber implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E210401	Shock from Patient Lead(s)	A0709	Device Sensing Problem		
									A071302	Inappropriate/Inadequate Shock/Stimulation		
							F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A0709	Device Sensing Problem		
									A071302	Inappropriate/Inadequate Shock/Stimulation		
							F1203	Life Threatening Illness or Injury	A0709	Device Sensing Problem		
									A071302	Inappropriate/Inadequate Shock/Stimulation		
16002	2024	All other reportable incidents	47265	Dual-chamber implantable pacemaker, rate-responsive	CT1685	Cardiac pacemakers	E011903	Syncope/Fainting	A070909	Oversensing	26/03/2024	26/03/2024
									A0712	Pacing Problem	26/03/2024	26/03/2024
									A0908	Incorrect, Inadequate or Imprecise Result or Readings	26/03/2024	26/03/2024
									A090801	Signal Artifact/Noise	26/03/2024	26/03/2024
							F08	Hospitalization or Prolonged Hospitalization	A070909	Oversensing	26/03/2024	26/03/2024
									A0712	Pacing Problem	26/03/2024	26/03/2024
									A0908	Incorrect, Inadequate or Imprecise Result or Readings	26/03/2024	26/03/2024
									A090801	Signal Artifact/Noise	26/03/2024	26/03/2024
							F0801	Intensive Care	A070909	Oversensing	26/03/2024	26/03/2024
									A0712	Pacing Problem	26/03/2024	26/03/2024
									A0908	Incorrect, Inadequate or Imprecise Result or Readings	26/03/2024	26/03/2024
									A090801	Signal Artifact/Noise	26/03/2024	26/03/2024
							F1001	Absence of Treatment	A070909	Oversensing	26/03/2024	26/03/2024
									A0712	Pacing Problem	26/03/2024	26/03/2024
									A0908	Incorrect, Inadequate or Imprecise Result or Readings	26/03/2024	26/03/2024
									A090801	Signal Artifact/Noise	26/03/2024	26/03/2024
							F1905	Device Revision or Replacement	A070909	Oversensing	26/03/2024	26/03/2024
									A0712	Pacing Problem	26/03/2024	26/03/2024
									A0908	Incorrect, Inadequate or Imprecise Result or Readings	26/03/2024	26/03/2024
									A090801	Signal Artifact/Noise	26/03/2024	26/03/2024
					CT257	Implantable cardiac pacemakers	E011903	Syncope/Fainting	A070909	Oversensing	26/03/2024	26/03/2024
									A0712	Pacing Problem	26/03/2024	26/03/2024
									A0908	Incorrect, Inadequate or Imprecise Result or Readings	26/03/2024	26/03/2024
									A090801	Signal Artifact/Noise	26/03/2024	26/03/2024
							F08	Hospitalization or Prolonged Hospitalization	A070909	Oversensing	26/03/2024	26/03/2024
									A0712	Pacing Problem	26/03/2024	26/03/2024
									A0908	Incorrect, Inadequate or Imprecise Result or Readings	26/03/2024	26/03/2024
									A090801	Signal Artifact/Noise	26/03/2024	26/03/2024
							F0801	Intensive Care	A070909	Oversensing	26/03/2024	26/03/2024
									A0712	Pacing Problem	26/03/2024	26/03/2024
									A0908	Incorrect, Inadequate or Imprecise Result or Readings	26/03/2024	26/03/2024
									A090801	Signal Artifact/Noise	26/03/2024	26/03/2024
							F1001	Absence of Treatment	A070909	Oversensing	26/03/2024	26/03/2024
									A0712	Pacing Problem	26/03/2024	26/03/2024
									A0908	Incorrect, Inadequate or Imprecise Result or Readings	26/03/2024	26/03/2024
									A090801	Signal Artifact/Noise	26/03/2024	26/03/2024
							F1905	Device Revision or Replacement	A070909	Oversensing	26/03/2024	26/03/2024
									A0712	Pacing Problem	26/03/2024	26/03/2024
									A0908	Incorrect, Inadequate or Imprecise Result or Readings	26/03/2024	26/03/2024
									A090801	Signal Artifact/Noise	26/03/2024	26/03/2024
					CT2866	Implantable cardiac electrical stimulators	E011903	Syncope/Fainting	A070909	Oversensing	26/03/2024	26/03/2024
									A0712	Pacing Problem	26/03/2024	26/03/2024

									A0908	Incorrect, Inadequate or Imprecise Result or Readings	26/03/2024	26/03/2024
									A090801	Signal Artifact/Noise	26/03/2024	26/03/2024
							F08	Hospitalization or Prolonged Hospitalization	A070909	Oversensing	26/03/2024	26/03/2024
									A0712	Pacing Problem	26/03/2024	26/03/2024
									A0908	Incorrect, Inadequate or Imprecise Result or Readings	26/03/2024	26/03/2024
									A090801	Signal Artifact/Noise	26/03/2024	26/03/2024
							F0801	Intensive Care	A070909	Oversensing	26/03/2024	26/03/2024
									A0712	Pacing Problem	26/03/2024	26/03/2024
									A0908	Incorrect, Inadequate or Imprecise Result or Readings	26/03/2024	26/03/2024
									A090801	Signal Artifact/Noise	26/03/2024	26/03/2024
							F1001	Absence of Treatment	A070909	Oversensing	26/03/2024	26/03/2024
									A0712	Pacing Problem	26/03/2024	26/03/2024
									A0908	Incorrect, Inadequate or Imprecise Result or Readings	26/03/2024	26/03/2024
									A090801	Signal Artifact/Noise	26/03/2024	26/03/2024
							F1905	Device Revision or Replacement	A070909	Oversensing	26/03/2024	26/03/2024
									A0712	Pacing Problem	26/03/2024	26/03/2024
									A0908	Incorrect, Inadequate or Imprecise Result or Readings	26/03/2024	26/03/2024
									A090801	Signal Artifact/Noise	26/03/2024	26/03/2024
16003	2024	All other reportable incidents	47265	Dual-chamber implantable pacemaker, rate-responsive	CT1685	Cardiac pacemakers	E0112	Dizziness	A070909	Oversensing		
									A071204	No Pacing		
							F1901	Additional Surgery	A070909	Oversensing		
									A071204	No Pacing		
					CT257	Implantable cardiac pacemakers	E0112	Dizziness	A070909	Oversensing		
									A071204	No Pacing		
							F1901	Additional Surgery	A070909	Oversensing		
									A071204	No Pacing		
					CT2866	Implantable cardiac electrical stimulators	E0112	Dizziness	A070909	Oversensing		
									A071204	No Pacing		
							F1901	Additional Surgery	A070909	Oversensing		
									A071204	No Pacing		
16004	2024	All other reportable incidents			CT1684	Cardiac pacemakers and associated devices	E060104	Bradycardia	A070101	Failure to Capture		
							F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A070101	Failure to Capture		
					CT257	Implantable cardiac pacemakers	E060104	Bradycardia	A070101	Failure to Capture		
							F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A070101	Failure to Capture		
16005	2024	All other reportable incidents	47263	Cardiac resynchronization therapy implantable pacemaker	CT1685	Cardiac pacemakers	E012207	Twitching	A071209	Extracardiac Stimulation	01/06/2024	01/06/2024
									A0908	Incorrect, Inadequate or Imprecise Result or Readings	01/06/2024	01/06/2024
							F08	Hospitalization or Prolonged Hospitalization	A071209	Extracardiac Stimulation	01/06/2024	01/06/2024
									A0908	Incorrect, Inadequate or Imprecise Result or Readings	01/06/2024	01/06/2024
							F1905	Device Revision or Replacement	A071209	Extracardiac Stimulation	01/06/2024	01/06/2024
									A0908	Incorrect, Inadequate or Imprecise Result or Readings	01/06/2024	01/06/2024
					CT257	Implantable cardiac pacemakers	E012207	Twitching	A071209	Extracardiac Stimulation	01/06/2024	01/06/2024
									A0908	Incorrect, Inadequate or Imprecise Result or Readings	01/06/2024	01/06/2024
							F08	Hospitalization or Prolonged Hospitalization	A071209	Extracardiac Stimulation	01/06/2024	01/06/2024
									A0908	Incorrect, Inadequate or Imprecise Result or Readings	01/06/2024	01/06/2024
							F1905	Device Revision or Replacement	A071209	Extracardiac Stimulation	01/06/2024	01/06/2024
									A0908	Incorrect, Inadequate or Imprecise Result or Readings	01/06/2024	01/06/2024
					CT2866	Implantable cardiac electrical stimulators	E012207	Twitching	A071209	Extracardiac Stimulation	01/06/2024	01/06/2024
									A0908	Incorrect, Inadequate or Imprecise Result or Readings	01/06/2024	01/06/2024
							F08	Hospitalization or Prolonged Hospitalization	A071209	Extracardiac Stimulation	01/06/2024	01/06/2024
									A0908	Incorrect, Inadequate or Imprecise Result or Readings	01/06/2024	01/06/2024
							F1905	Device Revision or Replacement	A071209	Extracardiac Stimulation	01/06/2024	01/06/2024
									A0908	Incorrect, Inadequate or Imprecise Result or Readings	01/06/2024	01/06/2024
16006	2024	All other reportable incidents			CT1684	Cardiac pacemakers and associated devices	E060102	Atrial Fibrillation	A27	Appropriate Device Problem Term/Code not Available		
							F11	Non-Serious Injury/ Illness/ Impairment	A27	Appropriate Device Problem Term/Code not Available		
					CT257	Implantable cardiac pacemakers	E060102	Atrial Fibrillation	A27	Appropriate Device Problem Term/Code not Available		
							F11	Non-Serious Injury/ Illness/ Impairment	A27	Appropriate Device Problem Term/Code not Available		
16007	2024	All other reportable incidents	17882	Physiologic-monitoring defibrillation system	CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem		
							F27	Problem Identified During Non-Clinical Procedure	A0705	Battery Problem		
16008	2024	All other reportable incidents	47265	Dual-chamber implantable pacemaker, rate-responsive	CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A05	Mechanical Problem	23/09/2024	23/09/2024
							F26	No Health Consequences or Impact	A05	Mechanical Problem	23/09/2024	23/09/2024
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A05	Mechanical Problem	23/09/2024	23/09/2024
							F26	No Health Consequences or Impact	A05	Mechanical Problem	23/09/2024	23/09/2024
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A05	Mechanical Problem	23/09/2024	23/09/2024
							F26	No Health Consequences or Impact	A05	Mechanical Problem	23/09/2024	23/09/2024
16009	2024	All other reportable incidents	47205	Cardiac pulse generator programmer	CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A0508	Noise, Audible		
									A070803	Failure to Power Up		
									A0719	Unexpected Shutdown		
							F26	No Health Consequences or Impact	A0508	Noise, Audible		
									A070803	Failure to Power Up		
									A0719	Unexpected Shutdown		
16010	2024	All other reportable incidents	17882	Physiologic-monitoring defibrillation system	CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A040103	Material Fragmentation		
							F27	Problem Identified During Non-Clinical Procedure	A040103	Material Fragmentation		
16011	2024	All other reportable incidents	17882	Physiologic-monitoring defibrillation system	CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A150105	Key or Button Unresponsive/not Working		
							F27	Problem Identified During Non-Clinical Procedure	A150105	Key or Button Unresponsive/not Working		

16012	2024	All other reportable incidents			CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070909	Oversensing		
									A072201	High Impedance		
									A090801	Signal Artifact/Noise		
							F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A070909	Oversensing		
									A072201	High Impedance		
									A090801	Signal Artifact/Noise		
							F2203	Imaging Required	A070909	Oversensing		
									A072201	High Impedance		
									A090801	Signal Artifact/Noise		
16013	2024	All other reportable incidents	17882	Physiologic-monitoring defibrillation system	CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A12	Connection Problem		
							F26	No Health Consequences or Impact	A12	Connection Problem		
16014	2024		47265	Dual-chamber implantable pacemaker, rate-responsive	CT1685	Cardiac pacemakers	E1906	Unspecified Infection	A25	No Apparent Adverse Event	26/03/2024	26/03/2024
							F26	No Health Consequences or Impact	A25	No Apparent Adverse Event	26/03/2024	26/03/2024
					CT257	Implantable cardiac pacemakers	E1906	Unspecified Infection	A25	No Apparent Adverse Event	26/03/2024	26/03/2024
							F26	No Health Consequences or Impact	A25	No Apparent Adverse Event	26/03/2024	26/03/2024
					CT2866	Implantable cardiac electrical stimulators	E1906	Unspecified Infection	A25	No Apparent Adverse Event	26/03/2024	26/03/2024
							F26	No Health Consequences or Impact	A25	No Apparent Adverse Event	26/03/2024	26/03/2024
16015	2024	All other reportable incidents	47205	Cardiac pulse generator programmer	CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A0902	Display or Visual Feedback Problem		
							F26	No Health Consequences or Impact	A0902	Display or Visual Feedback Problem		
16016	2024	All other reportable incidents	35852	Single-chamber implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E0112	Dizziness	A0709	Device Sensing Problem		
									A0712	Pacing Problem		
									A071302	Inappropriate/Inadequate Shock/Stimulation		
							E210401	Shock from Patient Lead(s)	A0709	Device Sensing Problem		
									A0712	Pacing Problem		
									A071302	Inappropriate/Inadequate Shock/Stimulation		
							E233001	Chest Pain	A0709	Device Sensing Problem		
									A0712	Pacing Problem		
									A071302	Inappropriate/Inadequate Shock/Stimulation		
							F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A0709	Device Sensing Problem		
									A0712	Pacing Problem		
									A071302	Inappropriate/Inadequate Shock/Stimulation		
							F1203	Life Threatening Illness or Injury	A0709	Device Sensing Problem		
									A0712	Pacing Problem		
									A071302	Inappropriate/Inadequate Shock/Stimulation		
16017	2024	Death	47270	Cardiac resynchronization therapy implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E060110	Ventricular Fibrillation	A0709	Device Sensing Problem		
									A071301	Failure to Deliver Shock/Stimulation		
									A090811	Unable to Obtain Readings		
							F02	Death	A0709	Device Sensing Problem		
									A071301	Failure to Deliver Shock/Stimulation		
									A090811	Unable to Obtain Readings		
16018	2024	Unanticipated serious deterioration in state of health	47265	Dual-chamber implantable pacemaker, rate-responsive	CT1685	Cardiac pacemakers	E0112	Dizziness	A0712	Pacing Problem		
							E060104	Bradycardia	A0712	Pacing Problem		
							E060106	Heart Block	A0712	Pacing Problem		
							F08	Hospitalization or Prolonged Hospitalization	A0712	Pacing Problem		
							F12	Serious Injury/ Illness/ Impairment	A0712	Pacing Problem		
							F2307	Reprogramming of Device	A0712	Pacing Problem		
					CT257	Implantable cardiac pacemakers	E0112	Dizziness	A0712	Pacing Problem		
							E060104	Bradycardia	A0712	Pacing Problem		
							E060106	Heart Block	A0712	Pacing Problem		
							F08	Hospitalization or Prolonged Hospitalization	A0712	Pacing Problem		
							F12	Serious Injury/ Illness/ Impairment	A0712	Pacing Problem		
							F2307	Reprogramming of Device	A0712	Pacing Problem		
					CT2866	Implantable cardiac electrical stimulators	E0112	Dizziness	A0712	Pacing Problem		
							E060104	Bradycardia	A0712	Pacing Problem		
							E060106	Heart Block	A0712	Pacing Problem		
							F08	Hospitalization or Prolonged Hospitalization	A0712	Pacing Problem		
							F12	Serious Injury/ Illness/ Impairment	A0712	Pacing Problem		
							F2307	Reprogramming of Device	A0712	Pacing Problem		
16019	2024	All other reportable incidents	35223	Endocardial/interventricular septal pacing lead	CT1684	Cardiac pacemakers and associated devices	E0618	Non Specific EKG/ECG Changes	A05	Mechanical Problem		
									A070101	Failure to Capture		
									A070909	Oversensing		
									A0712	Pacing Problem		
									A090801	Signal Artifact/Noise		
							F26	No Health Consequences or Impact	A05	Mechanical Problem		
									A070101	Failure to Capture		
									A070909	Oversensing		
									A0712	Pacing Problem		
									A090801	Signal Artifact/Noise		
16020	2024	All other reportable incidents	60789	Intracardiac pacemaker	CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A070104	Unstable Capture Threshold	02/10/2024	02/10/2024
									A1502	Positioning Problem	02/10/2024	02/10/2024
							F08	Hospitalization or Prolonged Hospitalization	A070104	Unstable Capture Threshold	02/10/2024	02/10/2024
									A1502	Positioning Problem	02/10/2024	02/10/2024
							F12	Serious Injury/ Illness/ Impairment	A070104	Unstable Capture Threshold	02/10/2024	02/10/2024
									A1502	Positioning Problem	02/10/2024	02/10/2024
							F1905	Device Revision or Replacement	A070104	Unstable Capture Threshold	02/10/2024	02/10/2024
									A1502	Positioning Problem	02/10/2024	02/10/2024
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A070104	Unstable Capture Threshold	02/10/2024	02/10/2024
									A1502	Positioning Problem	02/10/2024	02/10/2024
							F08	Hospitalization or Prolonged Hospitalization	A070104	Unstable Capture Threshold	02/10/2024	02/10/2024
									A1502	Positioning Problem	02/10/2024	02/10/2024
							F12	Serious Injury/ Illness/ Impairment	A070104	Unstable Capture Threshold	02/10/2024	02/10/2024
									A1502	Positioning Problem	02/10/2024	02/10/2024
							F1905	Device Revision or Replacement	A070104	Unstable Capture Threshold	02/10/2024	02/10/2024
									A1502	Positioning Problem	02/10/2024	02/10/2024
					CT2818	Catheterization/Cannulation/Fluid path devices	E2403	No Clinical Signs, Symptoms or Conditions	A070104	Unstable Capture Threshold	02/10/2024	02/10/2024
									A1502	Positioning Problem	02/10/2024	02/10/2024
							F08	Hospitalization or Prolonged Hospitalization	A070104	Unstable Capture Threshold	02/10/2024	02/10/2024
									A1502	Positioning Problem	02/10/2024	02/10/2024

							F12	Serious Injury/ Illness/ Impairment	A070104	Unstable Capture Threshold	02/10/2024	02/10/2024
									A1502	Positioning Problem	02/10/2024	02/10/2024
							F1905	Device Revision or Replacement	A070104	Unstable Capture Threshold	02/10/2024	02/10/2024
									A1502	Positioning Problem	02/10/2024	02/10/2024
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A070104	Unstable Capture Threshold	02/10/2024	02/10/2024
									A1502	Positioning Problem	02/10/2024	02/10/2024
							F08	Hospitalization or Prolonged Hospitalization	A070104	Unstable Capture Threshold	02/10/2024	02/10/2024
									A1502	Positioning Problem	02/10/2024	02/10/2024
							F12	Serious Injury/ Illness/ Impairment	A070104	Unstable Capture Threshold	02/10/2024	02/10/2024
									A1502	Positioning Problem	02/10/2024	02/10/2024
							F1905	Device Revision or Replacement	A070104	Unstable Capture Threshold	02/10/2024	02/10/2024
									A1502	Positioning Problem	02/10/2024	02/10/2024
16021	2024	All other reportable incidents	47265	Dual-chamber implantable pacemaker, rate-responsive	CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	30/04/2024	30/04/2024
							F1901	Additional Surgery	A24	No Apparent Device Problem	30/04/2024	30/04/2024
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	30/04/2024	30/04/2024
							F1901	Additional Surgery	A24	No Apparent Device Problem	30/04/2024	30/04/2024
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	30/04/2024	30/04/2024
							F1901	Additional Surgery	A24	No Apparent Device Problem	30/04/2024	30/04/2024
16022	2024		47265	Dual-chamber implantable pacemaker, rate-responsive	CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	07/03/2024	07/03/2024
							F1901	Additional Surgery	A24	No Apparent Device Problem	07/03/2024	07/03/2024
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	07/03/2024	07/03/2024
							F1901	Additional Surgery	A24	No Apparent Device Problem	07/03/2024	07/03/2024
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	07/03/2024	07/03/2024
							F1901	Additional Surgery	A24	No Apparent Device Problem	07/03/2024	07/03/2024
16023	2024	All other reportable incidents	47265	Dual-chamber implantable pacemaker, rate-responsive	CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A05	Mechanical Problem		
							F26	No Health Consequences or Impact	A05	Mechanical Problem		
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A05	Mechanical Problem		
							F26	No Health Consequences or Impact	A05	Mechanical Problem		
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A05	Mechanical Problem		
							F26	No Health Consequences or Impact	A05	Mechanical Problem		
16024	2024	All other reportable incidents	35852	Single-chamber implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E210401	Shock from Patient Lead(s)	A070909	Oversensing		
									A071302	Inappropriate/Inadequate Shock/Stimulation		
							F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A070909	Oversensing		
									A071302	Inappropriate/Inadequate Shock/Stimulation		
16025	2024	All other reportable incidents	37265	Dual-chamber implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	19/12/2024	19/12/2024
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	19/12/2024	19/12/2024
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	19/12/2024	19/12/2024
16026	2024		12913	Pacemaker, implantable	CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A0202	Defective Component		
									A090103	No Audible Prompt/Feedback		
									A1204	Failure to Disconnect		
							F26	No Health Consequences or Impact	A0202	Defective Component		
									A090103	No Audible Prompt/Feedback		
									A1204	Failure to Disconnect		
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A0202	Defective Component		
									A090103	No Audible Prompt/Feedback		
									A1204	Failure to Disconnect		
							F26	No Health Consequences or Impact	A0202	Defective Component		
									A090103	No Audible Prompt/Feedback		
									A1204	Failure to Disconnect		
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A0202	Defective Component		
									A090103	No Audible Prompt/Feedback		
									A1204	Failure to Disconnect		
							F26	No Health Consequences or Impact	A0202	Defective Component		
									A090103	No Audible Prompt/Feedback		
									A1204	Failure to Disconnect		
16027	2024	All other reportable incidents	17882	Physiologic-monitoring defibrillation system	CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A0719	Unexpected Shutdown		
							F26	No Health Consequences or Impact	A0719	Unexpected Shutdown		
16028	2024	All other reportable incidents	35854	Temporary cardiac pacing balloon catheter	CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A0712	Pacing Problem		
							F26	No Health Consequences or Impact	A0712	Pacing Problem		
16029	2024	All other reportable incidents	37265	Dual-chamber implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	28/10/2024	28/10/2024
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	28/10/2024	28/10/2024
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	28/10/2024	28/10/2024
16030	2024	All other reportable incidents	35852	Single-chamber implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	29/10/2024	29/10/2024
									A1304	Telemetry Discrepancy	29/10/2024	29/10/2024
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	29/10/2024	29/10/2024
									A1304	Telemetry Discrepancy	29/10/2024	29/10/2024
							F1903	Device Explantation	A070504	Premature Discharge of Battery	29/10/2024	29/10/2024
									A1304	Telemetry Discrepancy	29/10/2024	29/10/2024
16031	2024	All other reportable incidents	37265	Dual-chamber implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A05	Mechanical Problem	23/09/2024	23/09/2024
							F26	No Health Consequences or Impact	A05	Mechanical Problem	23/09/2024	23/09/2024
16032	2024	All other reportable incidents	35852	Single-chamber implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem	04/10/2024	04/10/2024
									A070504	Premature Discharge of Battery	04/10/2024	04/10/2024
							F08	Hospitalization or Prolonged Hospitalization	A0705	Battery Problem	04/10/2024	04/10/2024
									A070504	Premature Discharge of Battery	04/10/2024	04/10/2024
							F1905	Device Revision or Replacement	A0705	Battery Problem	04/10/2024	04/10/2024
									A070504	Premature Discharge of Battery	04/10/2024	04/10/2024
16033	2024	All other reportable incidents	17882	Physiologic-monitoring defibrillation system	CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A11	Computer Software Problem		
							F27	Problem Identified During Non-Clinical Procedure	A11	Computer Software Problem		
16034	2024	All other reportable incidents	17882	Physiologic-monitoring defibrillation system	CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A1603	Failure of Device to Self-Test		
							F27	Problem Identified During Non-Clinical Procedure	A1603	Failure of Device to Self-Test		
16035	2024		47270	Cardiac resynchronization therapy implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A25	No Apparent Adverse Event	07/10/2024	07/10/2024
							F1901	Additional Surgery	A25	No Apparent Adverse Event	07/10/2024	07/10/2024
16036	2024		47270	Cardiac resynchronization therapy implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E06	Heart	A0714	Unintended Electrical Shock		
							F05	Delay to Treatment/ Therapy	A0714	Unintended Electrical Shock		
16037	2024	All other reportable incidents	60789	Intracardiac pacemaker	CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A04	Material Integrity Problem	07/10/2024	07/10/2024
									A0504	Leak/Splash	07/10/2024	07/10/2024
									A1410	Difficult to Flush	07/10/2024	07/10/2024
							F26	No Health Consequences or Impact	A04	Material Integrity Problem	07/10/2024	07/10/2024

									A0504	Leak/Splash	07/10/2024	07/10/2024
									A1410	Difficult to Flush	07/10/2024	07/10/2024
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A04	Material Integrity Problem	07/10/2024	07/10/2024
									A0504	Leak/Splash	07/10/2024	07/10/2024
									A1410	Difficult to Flush	07/10/2024	07/10/2024
							F26	No Health Consequences or Impact	A04	Material Integrity Problem	07/10/2024	07/10/2024
									A0504	Leak/Splash	07/10/2024	07/10/2024
									A1410	Difficult to Flush	07/10/2024	07/10/2024
					CT2818	Catheterization/Cannulation/Fluid path devices	E2403	No Clinical Signs, Symptoms or Conditions	A04	Material Integrity Problem	07/10/2024	07/10/2024
									A0504	Leak/Splash	07/10/2024	07/10/2024
									A1410	Difficult to Flush	07/10/2024	07/10/2024
							F26	No Health Consequences or Impact	A04	Material Integrity Problem	07/10/2024	07/10/2024
									A0504	Leak/Splash	07/10/2024	07/10/2024
									A1410	Difficult to Flush	07/10/2024	07/10/2024
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A04	Material Integrity Problem	07/10/2024	07/10/2024
									A0504	Leak/Splash	07/10/2024	07/10/2024
									A1410	Difficult to Flush	07/10/2024	07/10/2024
							F26	No Health Consequences or Impact	A04	Material Integrity Problem	07/10/2024	07/10/2024
									A0504	Leak/Splash	07/10/2024	07/10/2024
									A1410	Difficult to Flush	07/10/2024	07/10/2024
16038	2024	All other reportable incidents			CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070101	Failure to Capture		
									A070909	Oversensing		
									A090801	Signal Artifact/Noise		
							F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A070101	Failure to Capture		
									A070909	Oversensing		
									A090801	Signal Artifact/Noise		
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A070101	Failure to Capture		
									A070909	Oversensing		
									A090801	Signal Artifact/Noise		
							F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A070101	Failure to Capture		
									A070909	Oversensing		
									A090801	Signal Artifact/Noise		
16039	2024		47263	Cardiac resynchronization therapy implantable pacemaker	CT1685	Cardiac pacemakers	E060106	Heart Block	A0203	Defective Device		
									A0705	Battery Problem		
									A071207	Pacing Inadequately		
							F0101	Therapeutic Response Decreased	A0203	Defective Device		
									A0705	Battery Problem		
									A071207	Pacing Inadequately		
							F08	Hospitalization or Prolonged Hospitalization	A0203	Defective Device		
									A0705	Battery Problem		
									A071207	Pacing Inadequately		
							F2301	Additional Device Required	A0203	Defective Device		
									A0705	Battery Problem		
									A071207	Pacing Inadequately		
					CT257	Implantable cardiac pacemakers	E060106	Heart Block	A0203	Defective Device		
									A0705	Battery Problem		
									A071207	Pacing Inadequately		
							F0101	Therapeutic Response Decreased	A0203	Defective Device		
									A0705	Battery Problem		
									A071207	Pacing Inadequately		
							F08	Hospitalization or Prolonged Hospitalization	A0203	Defective Device		
									A0705	Battery Problem		
									A071207	Pacing Inadequately		
							F2301	Additional Device Required	A0203	Defective Device		
									A0705	Battery Problem		
									A071207	Pacing Inadequately		
					CT2866	Implantable cardiac electrical stimulators	E060106	Heart Block	A0203	Defective Device		
									A0705	Battery Problem		
									A071207	Pacing Inadequately		
							F0101	Therapeutic Response Decreased	A0203	Defective Device		
									A0705	Battery Problem		
									A071207	Pacing Inadequately		
							F08	Hospitalization or Prolonged Hospitalization	A0203	Defective Device		
									A0705	Battery Problem		
									A071207	Pacing Inadequately		
							F2301	Additional Device Required	A0203	Defective Device		
									A0705	Battery Problem		
									A071207	Pacing Inadequately		
16040	2024	All other reportable incidents	47265	Dual-chamber implantable pacemaker, rate-responsive	CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	18/04/2024	18/04/2024
							F1901	Additional Surgery	A24	No Apparent Device Problem	18/04/2024	18/04/2024
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	18/04/2024	18/04/2024
							F1901	Additional Surgery	A24	No Apparent Device Problem	18/04/2024	18/04/2024
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	18/04/2024	18/04/2024
							F1901	Additional Surgery	A24	No Apparent Device Problem	18/04/2024	18/04/2024
16041	2024	All other reportable incidents	17882	Physiologic-monitoring defibrillation system	CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A11	Computer Software Problem		
							F27	Problem Identified During Non-Clinical Procedure	A11	Computer Software Problem		
16042	2024		47270	Cardiac resynchronization therapy implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E0601	Arrhythmia	A070603	Failure to Charge	07/10/2024	07/10/2024
									A071302	Inappropriate/Inadequate Shock/Stimulation	07/10/2024	07/10/2024
							F1901	Additional Surgery	A070603	Failure to Charge	07/10/2024	07/10/2024
									A071302	Inappropriate/Inadequate Shock/Stimulation	07/10/2024	07/10/2024
16043	2024	All other reportable incidents	47270	Cardiac resynchronization therapy implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
16044	2024	All other reportable incidents	47270	Cardiac resynchronization therapy implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A070909	Oversensing		
									A070910	Undersensing		
									A071204	No Pacing		
							F26	No Health Consequences or Impact	A070909	Oversensing		
									A070910	Undersensing		
									A071204	No Pacing		

16045	2024	All other reportable incidents	35852	Single-chamber implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E0112	Dizziness	A0709	Device Sensing Problem		
									A071302	Inappropriate/Inadequate Shock/Stimulation		
							E011902	Presyncope	A0709	Device Sensing Problem		
									A071302	Inappropriate/Inadequate Shock/Stimulation		
							F12	Serious Injury/ Illness/ Impairment	A0709	Device Sensing Problem		
									A071302	Inappropriate/Inadequate Shock/Stimulation		
							F2307	Reprogramming of Device	A0709	Device Sensing Problem		
									A071302	Inappropriate/Inadequate Shock/Stimulation		
16046	2024	All other reportable incidents	47270	Cardiac resynchronization therapy implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A070603	Failure to Charge	07/10/2024	07/10/2024
							F1901	Additional Surgery	A070603	Failure to Charge	07/10/2024	07/10/2024
16047	2024	All other reportable incidents	37265	Dual-chamber implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
16048	2024	All other reportable incidents	60789	Intracardiac pacemaker	CT1685	Cardiac pacemakers	E0112	Dizziness	A070101	Failure to Capture		
							F23	Unexpected Medical Intervention	A070101	Failure to Capture		
					CT257	Implantable cardiac pacemakers	E0112	Dizziness	A070101	Failure to Capture		
							F23	Unexpected Medical Intervention	A070101	Failure to Capture		
					CT2818	Catheterization/Cannulation/Fluid path devices	E0112	Dizziness	A070101	Failure to Capture		
							F23	Unexpected Medical Intervention	A070101	Failure to Capture		
					CT2866	Implantable cardiac electrical stimulators	E0112	Dizziness	A070101	Failure to Capture		
							F23	Unexpected Medical Intervention	A070101	Failure to Capture		
16049	2024	All other reportable incidents			CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A0712	Pacing Problem	08/10/2024	08/10/2024
							F1001	Absence of Treatment	A0712	Pacing Problem	08/10/2024	08/10/2024
							F1908	Prolonged Surgery	A0712	Pacing Problem	08/10/2024	08/10/2024
16050	2024	All other reportable incidents	47265	Dual-chamber implantable pacemaker, rate-responsive	CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem		
							F1903	Device Explantation	A0705	Battery Problem		
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem		
							F1903	Device Explantation	A0705	Battery Problem		
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem		
							F1903	Device Explantation	A0705	Battery Problem		
16051	2024	All other reportable incidents	60789	Intracardiac pacemaker	CT1685	Cardiac pacemakers	E0604	Cardiac Perforation	A24	No Apparent Device Problem		
							E0605	Cardiac Tamponade	A24	No Apparent Device Problem		
							E0619	Pericardial Effusion	A24	No Apparent Device Problem		
							E2321	Low Blood Pressure/ Hypotension	A24	No Apparent Device Problem		
							F1203	Life Threatening Illness or Injury	A24	No Apparent Device Problem		
					CT257	Implantable cardiac pacemakers	E0604	Cardiac Perforation	A24	No Apparent Device Problem		
							E0605	Cardiac Tamponade	A24	No Apparent Device Problem		
							E0619	Pericardial Effusion	A24	No Apparent Device Problem		
							E2321	Low Blood Pressure/ Hypotension	A24	No Apparent Device Problem		
							F1203	Life Threatening Illness or Injury	A24	No Apparent Device Problem		
					CT2818	Catheterization/Cannulation/Fluid path devices	E0604	Cardiac Perforation	A24	No Apparent Device Problem		
							E0605	Cardiac Tamponade	A24	No Apparent Device Problem		
							E0619	Pericardial Effusion	A24	No Apparent Device Problem		
							E2321	Low Blood Pressure/ Hypotension	A24	No Apparent Device Problem		
							F1203	Life Threatening Illness or Injury	A24	No Apparent Device Problem		
					CT2866	Implantable cardiac electrical stimulators	E0604	Cardiac Perforation	A24	No Apparent Device Problem		
							E0605	Cardiac Tamponade	A24	No Apparent Device Problem		
							E0619	Pericardial Effusion	A24	No Apparent Device Problem		
							E2321	Low Blood Pressure/ Hypotension	A24	No Apparent Device Problem		
							F1203	Life Threatening Illness or Injury	A24	No Apparent Device Problem		
16052	2024	All other reportable incidents	57968	Cardiovascular device introducer, steerable	CT1684	Cardiac pacemakers and associated devices	E0604	Cardiac Perforation	A24	No Apparent Device Problem		
							E0605	Cardiac Tamponade	A24	No Apparent Device Problem		
							E0619	Pericardial Effusion	A24	No Apparent Device Problem		
							E2321	Low Blood Pressure/ Hypotension	A24	No Apparent Device Problem		
							F1203	Life Threatening Illness or Injury	A24	No Apparent Device Problem		
16053	2024	All other reportable incidents	47265	Dual-chamber implantable pacemaker, rate-responsive	CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
16054	2024	All other reportable incidents	47263	Cardiac resynchronization therapy implantable pacemaker	CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A0908	Incorrect, Inadequate or Imprecise Result or Readings	07/10/2024	07/10/2024
							F08	Hospitalization or Prolonged Hospitalization	A0908	Incorrect, Inadequate or Imprecise Result or Readings	07/10/2024	07/10/2024
							F1905	Device Revision or Replacement	A0908	Incorrect, Inadequate or Imprecise Result or Readings	07/10/2024	07/10/2024
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A0908	Incorrect, Inadequate or Imprecise Result or Readings	07/10/2024	07/10/2024
							F08	Hospitalization or Prolonged Hospitalization	A0908	Incorrect, Inadequate or Imprecise Result or Readings	07/10/2024	07/10/2024
							F1905	Device Revision or Replacement	A0908	Incorrect, Inadequate or Imprecise Result or Readings	07/10/2024	07/10/2024
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A0908	Incorrect, Inadequate or Imprecise Result or Readings	07/10/2024	07/10/2024
							F08	Hospitalization or Prolonged Hospitalization	A0908	Incorrect, Inadequate or Imprecise Result or Readings	07/10/2024	07/10/2024
							F1905	Device Revision or Replacement	A0908	Incorrect, Inadequate or Imprecise Result or Readings	07/10/2024	07/10/2024
16055	2024	All other reportable incidents			CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070909	Oversensing		
									A072201	High Impedance		
							F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A070909	Oversensing		
									A072201	High Impedance		
16056	2024	All other reportable incidents	47270	Cardiac resynchronization therapy implantable	CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	05/11/2024	05/11/2024

				defibrillator					A0908	Incorrect, Inadequate or Imprecise Result or Readings	05/11/2024	05/11/2024
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	05/11/2024	05/11/2024
									A0908	Incorrect, Inadequate or Imprecise Result or Readings	05/11/2024	05/11/2024
							F1903	Device Explantation	A070504	Premature Discharge of Battery	05/11/2024	05/11/2024
									A0908	Incorrect, Inadequate or Imprecise Result or Readings	05/11/2024	05/11/2024
16057	2024	All other reportable incidents	47270	Cardiac resynchronization therapy implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem	05/12/2024	05/12/2024
									A070504	Premature Discharge of Battery	05/12/2024	05/12/2024
									A0908	Incorrect, Inadequate or Imprecise Result or Readings	05/12/2024	05/12/2024
							F08	Hospitalization or Prolonged Hospitalization	A0705	Battery Problem	05/12/2024	05/12/2024
									A070504	Premature Discharge of Battery	05/12/2024	05/12/2024
									A0908	Incorrect, Inadequate or Imprecise Result or Readings	05/12/2024	05/12/2024
							F1905	Device Revision or Replacement	A0705	Battery Problem	05/12/2024	05/12/2024
									A070504	Premature Discharge of Battery	05/12/2024	05/12/2024
									A0908	Incorrect, Inadequate or Imprecise Result or Readings	05/12/2024	05/12/2024
16058	2024	All other reportable incidents	47263	Cardiac resynchronization therapy implantable pacemaker	CT1685	Cardiac pacemakers	E0601	Arrhythmia	A070909	Oversensing		
									A0712	Pacing Problem		
									A072201	High Impedance		
									A090801	Signal Artifact/Noise		
							F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A070909	Oversensing		
									A0712	Pacing Problem		
									A072201	High Impedance		
									A090801	Signal Artifact/Noise		
					CT257	Implantable cardiac pacemakers	E0601	Arrhythmia	A070909	Oversensing		
									A0712	Pacing Problem		
									A072201	High Impedance		
									A090801	Signal Artifact/Noise		
							F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A070909	Oversensing		
									A0712	Pacing Problem		
									A072201	High Impedance		
									A090801	Signal Artifact/Noise		
					CT2866	Implantable cardiac electrical stimulators	E0601	Arrhythmia	A070909	Oversensing		
									A0712	Pacing Problem		
									A072201	High Impedance		
									A090801	Signal Artifact/Noise		
							F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A070909	Oversensing		
									A0712	Pacing Problem		
									A072201	High Impedance		
									A090801	Signal Artifact/Noise		
16059	2024	All other reportable incidents	47270	Cardiac resynchronization therapy implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E210401	Shock from Patient Lead(s)	A070504	Premature Discharge of Battery	15/12/2024	15/12/2024
									A070909	Oversensing	15/12/2024	15/12/2024
									A071302	Inappropriate/Inadequate Shock/Stimulation	15/12/2024	15/12/2024
									A0908	Incorrect, Inadequate or Imprecise Result or Readings	15/12/2024	15/12/2024
									A090801	Signal Artifact/Noise	15/12/2024	15/12/2024
									A150202	Malposition of Device	15/12/2024	15/12/2024
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	15/12/2024	15/12/2024
									A070909	Oversensing	15/12/2024	15/12/2024
									A071302	Inappropriate/Inadequate Shock/Stimulation	15/12/2024	15/12/2024
									A0908	Incorrect, Inadequate or Imprecise Result or Readings	15/12/2024	15/12/2024
									A090801	Signal Artifact/Noise	15/12/2024	15/12/2024
									A150202	Malposition of Device	15/12/2024	15/12/2024
							F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A070504	Premature Discharge of Battery	15/12/2024	15/12/2024
									A070909	Oversensing	15/12/2024	15/12/2024
									A071302	Inappropriate/Inadequate Shock/Stimulation	15/12/2024	15/12/2024
									A0908	Incorrect, Inadequate or Imprecise Result or Readings	15/12/2024	15/12/2024
									A090801	Signal Artifact/Noise	15/12/2024	15/12/2024
									A150202	Malposition of Device	15/12/2024	15/12/2024
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	15/12/2024	15/12/2024
									A070909	Oversensing	15/12/2024	15/12/2024
									A071302	Inappropriate/Inadequate Shock/Stimulation	15/12/2024	15/12/2024
									A0908	Incorrect, Inadequate or Imprecise Result or Readings	15/12/2024	15/12/2024
									A090801	Signal Artifact/Noise	15/12/2024	15/12/2024
									A150202	Malposition of Device	15/12/2024	15/12/2024
							F2203	Imaging Required	A070504	Premature Discharge of Battery	15/12/2024	15/12/2024
									A070909	Oversensing	15/12/2024	15/12/2024
									A071302	Inappropriate/Inadequate Shock/Stimulation	15/12/2024	15/12/2024
									A0908	Incorrect, Inadequate or Imprecise Result or Readings	15/12/2024	15/12/2024
									A090801	Signal Artifact/Noise	15/12/2024	15/12/2024
									A150202	Malposition of Device	15/12/2024	15/12/2024
16060	2024	All other reportable incidents	47205	Cardiac pulse generator programmer	CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A0709	Device Sensing Problem		
									A13	Communication or Transmission Problem		
							F26	No Health Consequences or Impact	A0709	Device Sensing Problem		
									A13	Communication or Transmission Problem		
16061	2024	All other reportable incidents	47263	Cardiac resynchronization therapy implantable pacemaker	CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	04/09/2024	04/09/2024
									A0908	Incorrect, Inadequate or Imprecise Result or Readings	04/09/2024	04/09/2024
									A1304	Telemetry Discrepancy	04/09/2024	04/09/2024
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	04/09/2024	04/09/2024
									A0908	Incorrect, Inadequate or Imprecise Result or Readings	04/09/2024	04/09/2024
									A1304	Telemetry Discrepancy	04/09/2024	04/09/2024

							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	04/09/2024	04/09/2024
									A0908	Incorrect, Inadequate or Imprecise Result or Readings	04/09/2024	04/09/2024
									A1304	Telemetry Discrepancy	04/09/2024	04/09/2024
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	04/09/2024	04/09/2024
									A0908	Incorrect, Inadequate or Imprecise Result or Readings	04/09/2024	04/09/2024
									A1304	Telemetry Discrepancy	04/09/2024	04/09/2024
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	04/09/2024	04/09/2024
									A0908	Incorrect, Inadequate or Imprecise Result or Readings	04/09/2024	04/09/2024
									A1304	Telemetry Discrepancy	04/09/2024	04/09/2024
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	04/09/2024	04/09/2024
									A0908	Incorrect, Inadequate or Imprecise Result or Readings	04/09/2024	04/09/2024
									A1304	Telemetry Discrepancy	04/09/2024	04/09/2024
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	04/09/2024	04/09/2024
									A0908	Incorrect, Inadequate or Imprecise Result or Readings	04/09/2024	04/09/2024
									A1304	Telemetry Discrepancy	04/09/2024	04/09/2024
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	04/09/2024	04/09/2024
									A0908	Incorrect, Inadequate or Imprecise Result or Readings	04/09/2024	04/09/2024
									A1304	Telemetry Discrepancy	04/09/2024	04/09/2024
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	04/09/2024	04/09/2024
									A0908	Incorrect, Inadequate or Imprecise Result or Readings	04/09/2024	04/09/2024
									A1304	Telemetry Discrepancy	04/09/2024	04/09/2024
16062	2024	All other reportable incidents	47270	Cardiac resynchronization therapy implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A070909	Oversensing		
									A071204	No Pacing		
									A071302	Inappropriate/Inadequate Shock/Stimulation		
							F12	Serious Injury/ Illness/ Impairment	A070909	Oversensing		
									A071204	No Pacing		
									A071302	Inappropriate/Inadequate Shock/Stimulation		
16063	2024	All other reportable incidents	47270	Cardiac resynchronization therapy implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	17/12/2024	17/12/2024
									A0908	Incorrect, Inadequate or Imprecise Result or Readings	17/12/2024	17/12/2024
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	17/12/2024	17/12/2024
									A0908	Incorrect, Inadequate or Imprecise Result or Readings	17/12/2024	17/12/2024
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	17/12/2024	17/12/2024
									A0908	Incorrect, Inadequate or Imprecise Result or Readings	17/12/2024	17/12/2024
16064	2024	All other reportable incidents	47265	Dual-chamber implantable pacemaker, rate-responsive	CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A04	Material Integrity Problem	08/10/2024	08/10/2024
									A12	Connection Problem	08/10/2024	08/10/2024
									A150206	Difficult to Insert	08/10/2024	08/10/2024
							F1908	Prolonged Surgery	A04	Material Integrity Problem	08/10/2024	08/10/2024
									A12	Connection Problem	08/10/2024	08/10/2024
									A150206	Difficult to Insert	08/10/2024	08/10/2024
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A04	Material Integrity Problem	08/10/2024	08/10/2024
									A12	Connection Problem	08/10/2024	08/10/2024
									A150206	Difficult to Insert	08/10/2024	08/10/2024
							F1908	Prolonged Surgery	A04	Material Integrity Problem	08/10/2024	08/10/2024
									A12	Connection Problem	08/10/2024	08/10/2024
									A150206	Difficult to Insert	08/10/2024	08/10/2024
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A04	Material Integrity Problem	08/10/2024	08/10/2024
									A12	Connection Problem	08/10/2024	08/10/2024
									A150206	Difficult to Insert	08/10/2024	08/10/2024
							F1908	Prolonged Surgery	A04	Material Integrity Problem	08/10/2024	08/10/2024
									A12	Connection Problem	08/10/2024	08/10/2024
									A150206	Difficult to Insert	08/10/2024	08/10/2024
16065	2024	All other reportable incidents	47267	Single-chamber implantable pacemaker, rate-responsive	CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A05	Mechanical Problem	10/09/2024	10/09/2024
							F26	No Health Consequences or Impact	A05	Mechanical Problem	10/09/2024	10/09/2024
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A05	Mechanical Problem	10/09/2024	10/09/2024
							F26	No Health Consequences or Impact	A05	Mechanical Problem	10/09/2024	10/09/2024
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A05	Mechanical Problem	10/09/2024	10/09/2024
							F26	No Health Consequences or Impact	A05	Mechanical Problem	10/09/2024	10/09/2024
16066	2024	All other reportable incidents	17882	Physiologic-monitoring defibrillation system	CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070802	Intermittent Loss of Power		
							F27	Problem Identified During Non-Clinical Procedure	A070802	Intermittent Loss of Power		
16067	2024		37265	Dual-chamber implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem		
							F1905	Device Revision or Replacement	A0705	Battery Problem		
							F23	Unexpected Medical Intervention	A0705	Battery Problem		
16068	2024	Unanticipated serious deterioration in state of health	47265	Dual-chamber implantable pacemaker, rate-responsive	CT1685	Cardiac pacemakers	E0119	Loss of Consciousness	A0712	Pacing Problem		
							E011903	Syncope/Fainting	A0712	Pacing Problem		
							E060106	Heart Block	A0712	Pacing Problem		
							E1603	Bone Fracture(s)	A0712	Pacing Problem		
							E2007	Fall	A0712	Pacing Problem		
							F08	Hospitalization or Prolonged Hospitalization	A0712	Pacing Problem		
							F1203	Life Threatening Illness or Injury	A0712	Pacing Problem		
					CT257	Implantable cardiac pacemakers	E0119	Loss of Consciousness	A0712	Pacing Problem		
							E011903	Syncope/Fainting	A0712	Pacing Problem		
							E060106	Heart Block	A0712	Pacing Problem		
							E1603	Bone Fracture(s)	A0712	Pacing Problem		
							E2007	Fall	A0712	Pacing Problem		
							F08	Hospitalization or Prolonged Hospitalization	A0712	Pacing Problem		
							F1203	Life Threatening Illness or Injury	A0712	Pacing Problem		
					CT2866	Implantable cardiac electrical stimulators	E0119	Loss of Consciousness	A0712	Pacing Problem		
							E011903	Syncope/Fainting	A0712	Pacing Problem		
							E060106	Heart Block	A0712	Pacing Problem		
							E1603	Bone Fracture(s)	A0712	Pacing Problem		

							E2007	Fall	A0712	Pacing Problem		
							F08	Hospitalization or Prolonged Hospitalization	A0712	Pacing Problem		
							F1203	Life Threatening Illness or Injury	A0712	Pacing Problem		
16069	2024	All other reportable incidents	47263	Cardiac resynchronization therapy implantable pacemaker	CT1685	Cardiac pacemakers	E011903	Syncope/Fainting	A070504	Premature Discharge of Battery	11/10/2024	11/10/2024
									A070909	Oversensing	11/10/2024	11/10/2024
									A0712	Pacing Problem	11/10/2024	11/10/2024
									A0908	Incorrect, Inadequate or Imprecise Result or Readings	11/10/2024	11/10/2024
									A090801	Signal Artifact/Noise	11/10/2024	11/10/2024
							E060101	Asystole	A070504	Premature Discharge of Battery	11/10/2024	11/10/2024
									A070909	Oversensing	11/10/2024	11/10/2024
									A0712	Pacing Problem	11/10/2024	11/10/2024
									A0908	Incorrect, Inadequate or Imprecise Result or Readings	11/10/2024	11/10/2024
									A090801	Signal Artifact/Noise	11/10/2024	11/10/2024
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	11/10/2024	11/10/2024
									A070909	Oversensing	11/10/2024	11/10/2024
									A0712	Pacing Problem	11/10/2024	11/10/2024
									A0908	Incorrect, Inadequate or Imprecise Result or Readings	11/10/2024	11/10/2024
									A090801	Signal Artifact/Noise	11/10/2024	11/10/2024
							F1001	Absence of Treatment	A070504	Premature Discharge of Battery	11/10/2024	11/10/2024
									A070909	Oversensing	11/10/2024	11/10/2024
									A0712	Pacing Problem	11/10/2024	11/10/2024
									A0908	Incorrect, Inadequate or Imprecise Result or Readings	11/10/2024	11/10/2024
									A090801	Signal Artifact/Noise	11/10/2024	11/10/2024
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	11/10/2024	11/10/2024
									A070909	Oversensing	11/10/2024	11/10/2024
									A0712	Pacing Problem	11/10/2024	11/10/2024
									A0908	Incorrect, Inadequate or Imprecise Result or Readings	11/10/2024	11/10/2024
									A090801	Signal Artifact/Noise	11/10/2024	11/10/2024
					CT257	Implantable cardiac pacemakers	E011903	Syncope/Fainting	A070504	Premature Discharge of Battery	11/10/2024	11/10/2024
									A070909	Oversensing	11/10/2024	11/10/2024
									A0712	Pacing Problem	11/10/2024	11/10/2024
									A0908	Incorrect, Inadequate or Imprecise Result or Readings	11/10/2024	11/10/2024
									A090801	Signal Artifact/Noise	11/10/2024	11/10/2024
							E060101	Asystole	A070504	Premature Discharge of Battery	11/10/2024	11/10/2024
									A070909	Oversensing	11/10/2024	11/10/2024
									A0712	Pacing Problem	11/10/2024	11/10/2024
									A0908	Incorrect, Inadequate or Imprecise Result or Readings	11/10/2024	11/10/2024
									A090801	Signal Artifact/Noise	11/10/2024	11/10/2024
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	11/10/2024	11/10/2024
									A070909	Oversensing	11/10/2024	11/10/2024
									A0712	Pacing Problem	11/10/2024	11/10/2024
									A0908	Incorrect, Inadequate or Imprecise Result or Readings	11/10/2024	11/10/2024
									A090801	Signal Artifact/Noise	11/10/2024	11/10/2024
							F1001	Absence of Treatment	A070504	Premature Discharge of Battery	11/10/2024	11/10/2024
									A070909	Oversensing	11/10/2024	11/10/2024
									A0712	Pacing Problem	11/10/2024	11/10/2024
									A0908	Incorrect, Inadequate or Imprecise Result or Readings	11/10/2024	11/10/2024
									A090801	Signal Artifact/Noise	11/10/2024	11/10/2024
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	11/10/2024	11/10/2024
									A070909	Oversensing	11/10/2024	11/10/2024
									A0712	Pacing Problem	11/10/2024	11/10/2024
									A0908	Incorrect, Inadequate or Imprecise Result or Readings	11/10/2024	11/10/2024
									A090801	Signal Artifact/Noise	11/10/2024	11/10/2024
					CT2866	Implantable cardiac electrical stimulators	E011903	Syncope/Fainting	A070504	Premature Discharge of Battery	11/10/2024	11/10/2024
									A070909	Oversensing	11/10/2024	11/10/2024
									A0712	Pacing Problem	11/10/2024	11/10/2024
									A0908	Incorrect, Inadequate or Imprecise Result or Readings	11/10/2024	11/10/2024
									A090801	Signal Artifact/Noise	11/10/2024	11/10/2024
							E060101	Asystole	A070504	Premature Discharge of Battery	11/10/2024	11/10/2024
									A070909	Oversensing	11/10/2024	11/10/2024
									A0712	Pacing Problem	11/10/2024	11/10/2024
									A0908	Incorrect, Inadequate or Imprecise Result or Readings	11/10/2024	11/10/2024
									A090801	Signal Artifact/Noise	11/10/2024	11/10/2024
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	11/10/2024	11/10/2024
									A070909	Oversensing	11/10/2024	11/10/2024
									A0712	Pacing Problem	11/10/2024	11/10/2024
									A0908	Incorrect, Inadequate or Imprecise Result or Readings	11/10/2024	11/10/2024
									A090801	Signal Artifact/Noise	11/10/2024	11/10/2024
							F1001	Absence of Treatment	A070504	Premature Discharge of Battery	11/10/2024	11/10/2024
									A070909	Oversensing	11/10/2024	11/10/2024
									A0712	Pacing Problem	11/10/2024	11/10/2024
									A0908	Incorrect, Inadequate or Imprecise Result or Readings	11/10/2024	11/10/2024
									A090801	Signal Artifact/Noise	11/10/2024	11/10/2024
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	11/10/2024	11/10/2024
									A070909	Oversensing	11/10/2024	11/10/2024
									A0712	Pacing Problem	11/10/2024	11/10/2024

									A0908	Incorrect, Inadequate or Imprecise Result or Readings	11/10/2024	11/10/2024
									A090801	Signal Artifact/Noise	11/10/2024	11/10/2024
16070	2024	All other reportable incidents	60789	Intracardiac pacemaker	CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A070102	High Capture Threshold		
							F08	Hospitalization or Prolonged Hospitalization	A070102	High Capture Threshold		
							F12	Serious Injury/ Illness/ Impairment	A070102	High Capture Threshold		
							F26	No Health Consequences or Impact	A070102	High Capture Threshold		
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A070102	High Capture Threshold		
							F08	Hospitalization or Prolonged Hospitalization	A070102	High Capture Threshold		
							F12	Serious Injury/ Illness/ Impairment	A070102	High Capture Threshold		
							F26	No Health Consequences or Impact	A070102	High Capture Threshold		
					CT2818	Catheterization/Cannulation/Fluid path devices	E2403	No Clinical Signs, Symptoms or Conditions	A070102	High Capture Threshold		
							F08	Hospitalization or Prolonged Hospitalization	A070102	High Capture Threshold		
							F12	Serious Injury/ Illness/ Impairment	A070102	High Capture Threshold		
							F26	No Health Consequences or Impact	A070102	High Capture Threshold		
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A070102	High Capture Threshold		
							F08	Hospitalization or Prolonged Hospitalization	A070102	High Capture Threshold		
							F12	Serious Injury/ Illness/ Impairment	A070102	High Capture Threshold		
							F26	No Health Consequences or Impact	A070102	High Capture Threshold		
16071	2024	All other reportable incidents	47265	Dual-chamber implantable pacemaker, rate-responsive	CT1685	Cardiac pacemakers	E0112	Dizziness	A070909	Oversensing		
									A0712	Pacing Problem		
									A072201	High Impedance		
									A090801	Signal Artifact/Noise		
							E0119	Loss of Consciousness	A070909	Oversensing		
									A0712	Pacing Problem		
									A072201	High Impedance		
									A090801	Signal Artifact/Noise		
							E011903	Syncope/Fainting	A070909	Oversensing		
									A0712	Pacing Problem		
									A072201	High Impedance		
									A090801	Signal Artifact/Noise		
							F1001	Absence of Treatment	A070909	Oversensing		
									A0712	Pacing Problem		
									A072201	High Impedance		
									A090801	Signal Artifact/Noise		
					CT257	Implantable cardiac pacemakers	E0112	Dizziness	A070909	Oversensing		
									A0712	Pacing Problem		
									A072201	High Impedance		
									A090801	Signal Artifact/Noise		
							E0119	Loss of Consciousness	A070909	Oversensing		
									A0712	Pacing Problem		
									A072201	High Impedance		
									A090801	Signal Artifact/Noise		
							E011903	Syncope/Fainting	A070909	Oversensing		
									A0712	Pacing Problem		
									A072201	High Impedance		
									A090801	Signal Artifact/Noise		
							F1001	Absence of Treatment	A070909	Oversensing		
									A0712	Pacing Problem		
									A072201	High Impedance		
									A090801	Signal Artifact/Noise		
					CT2866	Implantable cardiac electrical stimulators	E0112	Dizziness	A070909	Oversensing		
									A0712	Pacing Problem		
									A072201	High Impedance		
									A090801	Signal Artifact/Noise		
							E0119	Loss of Consciousness	A070909	Oversensing		
									A0712	Pacing Problem		
									A072201	High Impedance		
									A090801	Signal Artifact/Noise		
							E011903	Syncope/Fainting	A070909	Oversensing		
									A0712	Pacing Problem		
									A072201	High Impedance		
									A090801	Signal Artifact/Noise		
							F1001	Absence of Treatment	A070909	Oversensing		
									A0712	Pacing Problem		
									A072201	High Impedance		
									A090801	Signal Artifact/Noise		
16072	2024	All other reportable incidents	47265	Dual-chamber implantable pacemaker, rate-responsive	CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	16/09/2024	16/09/2024
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	16/09/2024	16/09/2024
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	16/09/2024	16/09/2024
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	16/09/2024	16/09/2024
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	16/09/2024	16/09/2024
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	16/09/2024	16/09/2024
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	16/09/2024	16/09/2024
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	16/09/2024	16/09/2024
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	16/09/2024	16/09/2024
16073	2024	All other reportable incidents	47263	Cardiac resynchronization therapy implantable pacemaker	CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	06/11/2024	06/11/2024
									A0908	Incorrect, Inadequate or Imprecise Result or Readings	06/11/2024	06/11/2024
									A1107	Data Problem	06/11/2024	06/11/2024
									A160503	Inappropriate or Unexpected Reset	06/11/2024	06/11/2024
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	06/11/2024	06/11/2024
									A0908	Incorrect, Inadequate or Imprecise Result or Readings	06/11/2024	06/11/2024
									A1107	Data Problem	06/11/2024	06/11/2024
									A160503	Inappropriate or Unexpected Reset	06/11/2024	06/11/2024
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	06/11/2024	06/11/2024
									A0908	Incorrect, Inadequate or Imprecise Result or Readings	06/11/2024	06/11/2024

									A1107	Data Problem	06/11/2024	06/11/2024
									A160503	Inappropriate or Unexpected Reset	06/11/2024	06/11/2024
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	06/11/2024	06/11/2024
									A0908	Incorrect, Inadequate or Imprecise Result or Readings	06/11/2024	06/11/2024
									A1107	Data Problem	06/11/2024	06/11/2024
									A160503	Inappropriate or Unexpected Reset	06/11/2024	06/11/2024
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	06/11/2024	06/11/2024
									A0908	Incorrect, Inadequate or Imprecise Result or Readings	06/11/2024	06/11/2024
									A1107	Data Problem	06/11/2024	06/11/2024
									A160503	Inappropriate or Unexpected Reset	06/11/2024	06/11/2024
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	06/11/2024	06/11/2024
									A0908	Incorrect, Inadequate or Imprecise Result or Readings	06/11/2024	06/11/2024
									A1107	Data Problem	06/11/2024	06/11/2024
									A160503	Inappropriate or Unexpected Reset	06/11/2024	06/11/2024
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	06/11/2024	06/11/2024
									A0908	Incorrect, Inadequate or Imprecise Result or Readings	06/11/2024	06/11/2024
									A1107	Data Problem	06/11/2024	06/11/2024
									A160503	Inappropriate or Unexpected Reset	06/11/2024	06/11/2024
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	06/11/2024	06/11/2024
									A0908	Incorrect, Inadequate or Imprecise Result or Readings	06/11/2024	06/11/2024
									A1107	Data Problem	06/11/2024	06/11/2024
									A160503	Inappropriate or Unexpected Reset	06/11/2024	06/11/2024
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	06/11/2024	06/11/2024
									A0908	Incorrect, Inadequate or Imprecise Result or Readings	06/11/2024	06/11/2024
									A1107	Data Problem	06/11/2024	06/11/2024
									A160503	Inappropriate or Unexpected Reset	06/11/2024	06/11/2024
16074	2024	All other reportable incidents	37265	Dual-chamber implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E060109	Tachycardia	A0713	Defibrillation/Stimulation Problem	17/10/2024	17/10/2024
									A0722	Impedance Problem	17/10/2024	17/10/2024
									A072202	Low Impedance	17/10/2024	17/10/2024
							E0602	Cardiac Arrest	A0713	Defibrillation/Stimulation Problem	17/10/2024	17/10/2024
									A0722	Impedance Problem	17/10/2024	17/10/2024
									A072202	Low Impedance	17/10/2024	17/10/2024
							F08	Hospitalization or Prolonged Hospitalization	A0713	Defibrillation/Stimulation Problem	17/10/2024	17/10/2024
									A0722	Impedance Problem	17/10/2024	17/10/2024
									A072202	Low Impedance	17/10/2024	17/10/2024
							F0801	Intensive Care	A0713	Defibrillation/Stimulation Problem	17/10/2024	17/10/2024
									A0722	Impedance Problem	17/10/2024	17/10/2024
									A072202	Low Impedance	17/10/2024	17/10/2024
							F1203	Life Threatening Illness or Injury	A0713	Defibrillation/Stimulation Problem	17/10/2024	17/10/2024
									A0722	Impedance Problem	17/10/2024	17/10/2024
									A072202	Low Impedance	17/10/2024	17/10/2024
							F1905	Device Revision or Replacement	A0713	Defibrillation/Stimulation Problem	17/10/2024	17/10/2024
									A0722	Impedance Problem	17/10/2024	17/10/2024
									A072202	Low Impedance	17/10/2024	17/10/2024
							F2306	Resuscitation	A0713	Defibrillation/Stimulation Problem	17/10/2024	17/10/2024
									A0722	Impedance Problem	17/10/2024	17/10/2024
									A072202	Low Impedance	17/10/2024	17/10/2024
							F2307	Reprogramming of Device	A0713	Defibrillation/Stimulation Problem	17/10/2024	17/10/2024
									A0722	Impedance Problem	17/10/2024	17/10/2024
									A072202	Low Impedance	17/10/2024	17/10/2024
16075	2024	All other reportable incidents	47265	Dual-chamber implantable pacemaker, rate-responsive	CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A0908	Incorrect, Inadequate or Imprecise Result or Readings	07/10/2024	07/10/2024
							F08	Hospitalization or Prolonged Hospitalization	A0908	Incorrect, Inadequate or Imprecise Result or Readings	07/10/2024	07/10/2024
							F1905	Device Revision or Replacement	A0908	Incorrect, Inadequate or Imprecise Result or Readings	07/10/2024	07/10/2024
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A0908	Incorrect, Inadequate or Imprecise Result or Readings	07/10/2024	07/10/2024
							F08	Hospitalization or Prolonged Hospitalization	A0908	Incorrect, Inadequate or Imprecise Result or Readings	07/10/2024	07/10/2024
							F1905	Device Revision or Replacement	A0908	Incorrect, Inadequate or Imprecise Result or Readings	07/10/2024	07/10/2024
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A0908	Incorrect, Inadequate or Imprecise Result or Readings	07/10/2024	07/10/2024
							F08	Hospitalization or Prolonged Hospitalization	A0908	Incorrect, Inadequate or Imprecise Result or Readings	07/10/2024	07/10/2024
							F1905	Device Revision or Replacement	A0908	Incorrect, Inadequate or Imprecise Result or Readings	07/10/2024	07/10/2024
16076	2024	All other reportable incidents	47270	Cardiac resynchronization therapy implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A072001	Electromagnetic Interference		
									A090801	Signal Artifact/Noise		
							F26	No Health Consequences or Impact	A072001	Electromagnetic Interference		
									A090801	Signal Artifact/Noise		
16077	2024	All other reportable incidents	47263	Cardiac resynchronization therapy implantable pacemaker	CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A0908	Incorrect, Inadequate or Imprecise Result or Readings		
							F26	No Health Consequences or Impact	A0908	Incorrect, Inadequate or Imprecise Result or Readings		
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A0908	Incorrect, Inadequate or Imprecise Result or Readings		
							F26	No Health Consequences or Impact	A0908	Incorrect, Inadequate or Imprecise Result or Readings		
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A0908	Incorrect, Inadequate or Imprecise Result or Readings		
							F26	No Health Consequences or Impact	A0908	Incorrect, Inadequate or Imprecise Result or Readings		
16078	2024	All other reportable incidents	47263	Cardiac resynchronization therapy implantable pacemaker	CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A0908	Incorrect, Inadequate or Imprecise Result or Readings	09/10/2024	09/10/2024

							F08	Hospitalization or Prolonged Hospitalization	A0908	Incorrect, Inadequate or Imprecise Result or Readings	09/10/2024	09/10/2024
							F1905	Device Revision or Replacement	A0908	Incorrect, Inadequate or Imprecise Result or Readings	09/10/2024	09/10/2024
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A0908	Incorrect, Inadequate or Imprecise Result or Readings	09/10/2024	09/10/2024
							F08	Hospitalization or Prolonged Hospitalization	A0908	Incorrect, Inadequate or Imprecise Result or Readings	09/10/2024	09/10/2024
							F1905	Device Revision or Replacement	A0908	Incorrect, Inadequate or Imprecise Result or Readings	09/10/2024	09/10/2024
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A0908	Incorrect, Inadequate or Imprecise Result or Readings	09/10/2024	09/10/2024
							F08	Hospitalization or Prolonged Hospitalization	A0908	Incorrect, Inadequate or Imprecise Result or Readings	09/10/2024	09/10/2024
							F1905	Device Revision or Replacement	A0908	Incorrect, Inadequate or Imprecise Result or Readings	09/10/2024	09/10/2024
16079	2024	All other reportable incidents	47265	Dual-chamber implantable pacemaker, rate-responsive	CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	08/10/2024	08/10/2024
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	08/10/2024	08/10/2024
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	08/10/2024	08/10/2024
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	08/10/2024	08/10/2024
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	08/10/2024	08/10/2024
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	08/10/2024	08/10/2024
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	08/10/2024	08/10/2024
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	08/10/2024	08/10/2024
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	08/10/2024	08/10/2024
16080	2024	All other reportable incidents	47265	Dual-chamber implantable pacemaker, rate-responsive	CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	10/10/2024	10/10/2024
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	10/10/2024	10/10/2024
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	10/10/2024	10/10/2024
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	10/10/2024	10/10/2024
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	10/10/2024	10/10/2024
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	10/10/2024	10/10/2024
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	10/10/2024	10/10/2024
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	10/10/2024	10/10/2024
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	10/10/2024	10/10/2024
16081	2024		47265	Dual-chamber implantable pacemaker, rate-responsive	CT1685	Cardiac pacemakers	E233001	Chest Pain	A01	Patient Device Interaction Problem		
							F11	Non-Serious Injury/ Illness/ Impairment	A01	Patient Device Interaction Problem		
					CT257	Implantable cardiac pacemakers	E233001	Chest Pain	A01	Patient Device Interaction Problem		
							F11	Non-Serious Injury/ Illness/ Impairment	A01	Patient Device Interaction Problem		
					CT2866	Implantable cardiac electrical stimulators	E233001	Chest Pain	A01	Patient Device Interaction Problem		
							F11	Non-Serious Injury/ Illness/ Impairment	A01	Patient Device Interaction Problem		
16082	2024	All other reportable incidents	37265	Dual-chamber implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A070101	Failure to Capture		
									A070102	High Capture Threshold		
									A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
									A072201	High Impedance		
									A0908	Incorrect, Inadequate or Imprecise Result or Readings		
							F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A070101	Failure to Capture		
									A070102	High Capture Threshold		
									A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
									A072201	High Impedance		
									A0908	Incorrect, Inadequate or Imprecise Result or Readings		
16083	2024	All other reportable incidents	17882	Physiologic-monitoring defibrillation system	CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A1103	Program or Algorithm Execution Problem		
							F27	Problem Identified During Non-Clinical Procedure	A1103	Program or Algorithm Execution Problem		
16084	2024	All other reportable incidents	17882	Physiologic-monitoring defibrillation system	CT1684	Cardiac pacemakers and associated devices	E2401	Insufficient Information	A26	Insufficient Device Problem Information		
							F26	No Health Consequences or Impact	A26	Insufficient Device Problem Information		
16085	2024	All other reportable incidents	37265	Dual-chamber implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A160601	Premature Elective Replacement Indicator		
							F1903	Device Explantation	A160601	Premature Elective Replacement Indicator		
16086	2024	All other reportable incidents	35852	Single-chamber implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A160601	Premature Elective Replacement Indicator		
							F08	Hospitalization or Prolonged Hospitalization	A160601	Premature Elective Replacement Indicator		
16087	2024	All other reportable incidents	47270	Cardiac resynchronization therapy implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F12	Serious Injury/ Illness/ Impairment	A070504	Premature Discharge of Battery		
							F2307	Reprogramming of Device	A070504	Premature Discharge of Battery		
16088	2024	All other reportable incidents	47265	Dual-chamber implantable pacemaker, rate-responsive	CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A071204	No Pacing	03/09/2024	03/09/2024
							F1901	Additional Surgery	A071204	No Pacing	03/09/2024	03/09/2024
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A071204	No Pacing	03/09/2024	03/09/2024
							F1901	Additional Surgery	A071204	No Pacing	03/09/2024	03/09/2024
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A071204	No Pacing	03/09/2024	03/09/2024
							F1901	Additional Surgery	A071204	No Pacing	03/09/2024	03/09/2024
16089	2024	All other reportable incidents	47270	Cardiac resynchronization therapy implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A070909	Oversensing		
									A072001	Electromagnetic Interference		
									A090801	Signal Artifact/Noise		
									A160105	False Alarm		
							F26	No Health Consequences or Impact	A070909	Oversensing		
									A072001	Electromagnetic Interference		
									A090801	Signal Artifact/Noise		
									A160105	False Alarm		
16090	2024		47265	Dual-chamber implantable pacemaker, rate-responsive	CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	05/10/2024	05/10/2024
							F1901	Additional Surgery	A24	No Apparent Device Problem	05/10/2024	05/10/2024
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	05/10/2024	05/10/2024
							F1901	Additional Surgery	A24	No Apparent Device Problem	05/10/2024	05/10/2024
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	05/10/2024	05/10/2024
							F1901	Additional Surgery	A24	No Apparent Device Problem	05/10/2024	05/10/2024
16091	2024	All other reportable incidents	35852	Single-chamber implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	01/08/2024	01/08/2024
									A0908	Incorrect, Inadequate or Imprecise Result or Readings	01/08/2024	01/08/2024
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	01/08/2024	01/08/2024
									A0908	Incorrect, Inadequate or Imprecise Result or Readings	01/08/2024	01/08/2024

							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	01/08/2024	01/08/2024
									A0908	Incorrect, Inadequate or Imprecise Result or Readings	01/08/2024	01/08/2024
16092	2024	All other reportable incidents	47265	Dual-chamber implantable pacemaker, rate-responsive	CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem	12/03/2024	12/03/2024
									A070504	Premature Discharge of Battery	12/03/2024	12/03/2024
							F08	Hospitalization or Prolonged Hospitalization	A0705	Battery Problem	12/03/2024	12/03/2024
									A070504	Premature Discharge of Battery	12/03/2024	12/03/2024
							F1905	Device Revision or Replacement	A0705	Battery Problem	12/03/2024	12/03/2024
									A070504	Premature Discharge of Battery	12/03/2024	12/03/2024
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem	12/03/2024	12/03/2024
									A070504	Premature Discharge of Battery	12/03/2024	12/03/2024
							F08	Hospitalization or Prolonged Hospitalization	A0705	Battery Problem	12/03/2024	12/03/2024
									A070504	Premature Discharge of Battery	12/03/2024	12/03/2024
							F1905	Device Revision or Replacement	A0705	Battery Problem	12/03/2024	12/03/2024
									A070504	Premature Discharge of Battery	12/03/2024	12/03/2024
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem	12/03/2024	12/03/2024
									A070504	Premature Discharge of Battery	12/03/2024	12/03/2024
							F08	Hospitalization or Prolonged Hospitalization	A0705	Battery Problem	12/03/2024	12/03/2024
									A070504	Premature Discharge of Battery	12/03/2024	12/03/2024
							F1905	Device Revision or Replacement	A0705	Battery Problem	12/03/2024	12/03/2024
									A070504	Premature Discharge of Battery	12/03/2024	12/03/2024
16093	2024	All other reportable incidents	47265	Dual-chamber implantable pacemaker, rate-responsive	CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A0908	Incorrect, Inadequate or Imprecise Result or Readings	17/10/2024	17/10/2024
							F08	Hospitalization or Prolonged Hospitalization	A0908	Incorrect, Inadequate or Imprecise Result or Readings	17/10/2024	17/10/2024
							F1905	Device Revision or Replacement	A0908	Incorrect, Inadequate or Imprecise Result or Readings	17/10/2024	17/10/2024
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A0908	Incorrect, Inadequate or Imprecise Result or Readings	17/10/2024	17/10/2024
							F08	Hospitalization or Prolonged Hospitalization	A0908	Incorrect, Inadequate or Imprecise Result or Readings	17/10/2024	17/10/2024
							F1905	Device Revision or Replacement	A0908	Incorrect, Inadequate or Imprecise Result or Readings	17/10/2024	17/10/2024
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A0908	Incorrect, Inadequate or Imprecise Result or Readings	17/10/2024	17/10/2024
							F08	Hospitalization or Prolonged Hospitalization	A0908	Incorrect, Inadequate or Imprecise Result or Readings	17/10/2024	17/10/2024
							F1905	Device Revision or Replacement	A0908	Incorrect, Inadequate or Imprecise Result or Readings	17/10/2024	17/10/2024
16094	2024	All other reportable incidents	35852	Single-chamber implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A072201	High Impedance	18/11/2024	18/11/2024
							F2203	Imaging Required	A072201	High Impedance	18/11/2024	18/11/2024
16095	2024		47265	Dual-chamber implantable pacemaker, rate-responsive	CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	12/09/2024	12/09/2024
							F1901	Additional Surgery	A24	No Apparent Device Problem	12/09/2024	12/09/2024
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	12/09/2024	12/09/2024
							F1901	Additional Surgery	A24	No Apparent Device Problem	12/09/2024	12/09/2024
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	12/09/2024	12/09/2024
							F1901	Additional Surgery	A24	No Apparent Device Problem	12/09/2024	12/09/2024
16096	2024	All other reportable incidents	35852	Single-chamber implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	15/10/2024	15/10/2024
									A0908	Incorrect, Inadequate or Imprecise Result or Readings	15/10/2024	15/10/2024
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	15/10/2024	15/10/2024
									A0908	Incorrect, Inadequate or Imprecise Result or Readings	15/10/2024	15/10/2024
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	15/10/2024	15/10/2024
									A0908	Incorrect, Inadequate or Imprecise Result or Readings	15/10/2024	15/10/2024
16097	2024	All other reportable incidents	47265	Dual-chamber implantable pacemaker, rate-responsive	CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem		
							F26	No Health Consequences or Impact	A24	No Apparent Device Problem		
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem		
							F26	No Health Consequences or Impact	A24	No Apparent Device Problem		
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem		
							F26	No Health Consequences or Impact	A24	No Apparent Device Problem		
16098	2024	All other reportable incidents			CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070909	Oversensing		
							F10	Inadequate/inappropriate Treatment or Diagnostic Exposure	A070909	Oversensing		
16099	2024	All other reportable incidents	47270	Cardiac resynchronization therapy implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem		
									A1107	Data Problem		
							F26	No Health Consequences or Impact	A0705	Battery Problem		
									A1107	Data Problem		
16100	2024		47269	Transtelephonic implantable pacemaker analysis system	CT1684	Cardiac pacemakers and associated devices	E2401	Insufficient Information	A26	Insufficient Device Problem Information		
							F26	No Health Consequences or Impact	A26	Insufficient Device Problem Information		
16101	2024	All other reportable incidents	47265	Dual-chamber implantable pacemaker, rate-responsive	CT1685	Cardiac pacemakers	E0601	Arrhythmia	A0712	Pacing Problem		
							F11	Non-Serious Injury/ Illness/ Impairment	A0712	Pacing Problem		
					CT257	Implantable cardiac pacemakers	E0601	Arrhythmia	A0712	Pacing Problem		
							F11	Non-Serious Injury/ Illness/ Impairment	A0712	Pacing Problem		
					CT2866	Implantable cardiac electrical stimulators	E0601	Arrhythmia	A0712	Pacing Problem		
							F11	Non-Serious Injury/ Illness/ Impairment	A0712	Pacing Problem		
16102	2024	All other reportable incidents	47265	Dual-chamber implantable pacemaker, rate-responsive	CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem	09/10/2024	09/10/2024
							F08	Hospitalization or Prolonged Hospitalization	A0705	Battery Problem	09/10/2024	09/10/2024
							F1905	Device Revision or Replacement	A0705	Battery Problem	09/10/2024	09/10/2024
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem	09/10/2024	09/10/2024
							F08	Hospitalization or Prolonged Hospitalization	A0705	Battery Problem	09/10/2024	09/10/2024
							F1905	Device Revision or Replacement	A0705	Battery Problem	09/10/2024	09/10/2024
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem	09/10/2024	09/10/2024
							F08	Hospitalization or Prolonged Hospitalization	A0705	Battery Problem	09/10/2024	09/10/2024
							F1905	Device Revision or Replacement	A0705	Battery Problem	09/10/2024	09/10/2024
16103	2024		47270	Cardiac resynchronization therapy implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E011903	Syncope/Fainting	A071301	Failure to Deliver Shock/Stimulation	01/01/2024	31/12/2024
									A160503	Inappropriate or Unexpected Reset	01/01/2024	31/12/2024
							E0601	Arrhythmia	A071301	Failure to Deliver Shock/Stimulation	01/01/2024	31/12/2024
									A160503	Inappropriate or Unexpected Reset	01/01/2024	31/12/2024
							F12	Serious Injury/ Illness/ Impairment	A071301	Failure to Deliver Shock/Stimulation	01/01/2024	31/12/2024
									A160503	Inappropriate or Unexpected Reset	01/01/2024	31/12/2024

16104	2024	All other reportable incidents			CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070909	Oversensing		
							F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A070909	Oversensing		
16105	2024	All other reportable incidents	35852	Single-chamber implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A070909	Oversensing		
									A070910	Undersensing		
									A0712	Pacing Problem		
							F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A070909	Oversensing		
									A070910	Undersensing		
									A0712	Pacing Problem		
16106	2024	All other reportable incidents	47265	Dual-chamber implantable pacemaker, rate-responsive	CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A0712	Pacing Problem		
							F12	Serious Injury/ Illness/ Impairment	A0712	Pacing Problem		
							F2307	Reprogramming of Device	A0712	Pacing Problem		
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A0712	Pacing Problem		
							F12	Serious Injury/ Illness/ Impairment	A0712	Pacing Problem		
							F2307	Reprogramming of Device	A0712	Pacing Problem		
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A0712	Pacing Problem		
							F12	Serious Injury/ Illness/ Impairment	A0712	Pacing Problem		
							F2307	Reprogramming of Device	A0712	Pacing Problem		
16107	2024	All other reportable incidents	17882	Physiologic-monitoring defibrillation system	CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A150105	Key or Button Unresponsive/not Working		
							F27	Problem Identified During Non-Clinical Procedure	A150105	Key or Button Unresponsive/not Working		
16108	2024	All other reportable incidents	47265	Dual-chamber implantable pacemaker, rate-responsive	CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	18/10/2024	18/10/2024
							F1901	Additional Surgery	A24	No Apparent Device Problem	18/10/2024	18/10/2024
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	18/10/2024	18/10/2024
							F1901	Additional Surgery	A24	No Apparent Device Problem	18/10/2024	18/10/2024
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	18/10/2024	18/10/2024
							F1901	Additional Surgery	A24	No Apparent Device Problem	18/10/2024	18/10/2024
16109	2024	All other reportable incidents	47265	Dual-chamber implantable pacemaker, rate-responsive	CT1685	Cardiac pacemakers	E0717	Dyspnea	A0908	Incorrect, Inadequate or Imprecise Result or Readings	21/10/2024	21/10/2024
							F08	Hospitalization or Prolonged Hospitalization	A0908	Incorrect, Inadequate or Imprecise Result or Readings	21/10/2024	21/10/2024
							F1905	Device Revision or Replacement	A0908	Incorrect, Inadequate or Imprecise Result or Readings	21/10/2024	21/10/2024
					CT257	Implantable cardiac pacemakers	E0717	Dyspnea	A0908	Incorrect, Inadequate or Imprecise Result or Readings	21/10/2024	21/10/2024
							F08	Hospitalization or Prolonged Hospitalization	A0908	Incorrect, Inadequate or Imprecise Result or Readings	21/10/2024	21/10/2024
							F1905	Device Revision or Replacement	A0908	Incorrect, Inadequate or Imprecise Result or Readings	21/10/2024	21/10/2024
					CT2866	Implantable cardiac electrical stimulators	E0717	Dyspnea	A0908	Incorrect, Inadequate or Imprecise Result or Readings	21/10/2024	21/10/2024
							F08	Hospitalization or Prolonged Hospitalization	A0908	Incorrect, Inadequate or Imprecise Result or Readings	21/10/2024	21/10/2024
							F1905	Device Revision or Replacement	A0908	Incorrect, Inadequate or Imprecise Result or Readings	21/10/2024	21/10/2024
16110	2024	All other reportable incidents	17882	Physiologic-monitoring defibrillation system	CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A0713	Defibrillation/Stimulation Problem		
							F27	Problem Identified During Non-Clinical Procedure	A0713	Defibrillation/Stimulation Problem		
16111	2024	Unanticipated serious deterioration in state of health	37265	Dual-chamber implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E0610	Endocarditis	A24	No Apparent Device Problem	17/10/2024	17/10/2024
							E062104	Tricuspid Valve Insufficiency/ Regurgitation	A24	No Apparent Device Problem	17/10/2024	17/10/2024
							E1901	Bacterial Infection	A24	No Apparent Device Problem	17/10/2024	17/10/2024
							F08	Hospitalization or Prolonged Hospitalization	A24	No Apparent Device Problem	17/10/2024	17/10/2024
							F12	Serious Injury/ Illness/ Impairment	A24	No Apparent Device Problem	17/10/2024	17/10/2024
							F1903	Device Explantation	A24	No Apparent Device Problem	17/10/2024	17/10/2024
16112	2024		47265	Dual-chamber implantable pacemaker, rate-responsive	CT1685	Cardiac pacemakers	E0717	Dyspnea	A051206	Unintended System Motion		
							F1903	Device Explantation	A051206	Unintended System Motion		
					CT257	Implantable cardiac pacemakers	E0717	Dyspnea	A051206	Unintended System Motion		
							F1903	Device Explantation	A051206	Unintended System Motion		
					CT2866	Implantable cardiac electrical stimulators	E0717	Dyspnea	A051206	Unintended System Motion		
							F1903	Device Explantation	A051206	Unintended System Motion		
16113	2024		35974	Cardiac electrophysiology stimulation system	CT1684	Cardiac pacemakers and associated devices	F26	No Health Consequences or Impact	A0705	Battery Problem		
16114	2024	All other reportable incidents	47265	Dual-chamber implantable pacemaker, rate-responsive	CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A0203	Defective Device	20/05/2024	20/05/2024
							F24	Insufficient Health Impact Information	A0203	Defective Device	20/05/2024	20/05/2024
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A0203	Defective Device	20/05/2024	20/05/2024
							F24	Insufficient Health Impact Information	A0203	Defective Device	20/05/2024	20/05/2024
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A0203	Defective Device	20/05/2024	20/05/2024
							F24	Insufficient Health Impact Information	A0203	Defective Device	20/05/2024	20/05/2024
16115	2024	All other reportable incidents	47265	Dual-chamber implantable pacemaker, rate-responsive	CT1685	Cardiac pacemakers	E0717	Dyspnea	A071102	Failure to Interrogate	23/10/2024	23/10/2024
									A0723	Unexpected Mode Switch	23/10/2024	23/10/2024
							E2403	No Clinical Signs, Symptoms or Conditions	A071102	Failure to Interrogate	23/10/2024	23/10/2024
									A0723	Unexpected Mode Switch	23/10/2024	23/10/2024
							F1901	Additional Surgery	A071102	Failure to Interrogate	23/10/2024	23/10/2024
									A0723	Unexpected Mode Switch	23/10/2024	23/10/2024
					CT257	Implantable cardiac pacemakers	E0717	Dyspnea	A071102	Failure to Interrogate	23/10/2024	23/10/2024
									A0723	Unexpected Mode Switch	23/10/2024	23/10/2024
							E2403	No Clinical Signs, Symptoms or Conditions	A071102	Failure to Interrogate	23/10/2024	23/10/2024
									A0723	Unexpected Mode Switch	23/10/2024	23/10/2024
							F1901	Additional Surgery	A071102	Failure to Interrogate	23/10/2024	23/10/2024
									A0723	Unexpected Mode Switch	23/10/2024	23/10/2024
					CT2866	Implantable cardiac electrical stimulators	E0717	Dyspnea	A071102	Failure to Interrogate	23/10/2024	23/10/2024
									A0723	Unexpected Mode Switch	23/10/2024	23/10/2024
							E2403	No Clinical Signs, Symptoms or Conditions	A071102	Failure to Interrogate	23/10/2024	23/10/2024
									A0723	Unexpected Mode Switch	23/10/2024	23/10/2024
							F1901	Additional Surgery	A071102	Failure to Interrogate	23/10/2024	23/10/2024
									A0723	Unexpected Mode Switch	23/10/2024	23/10/2024
16116	2024	All other reportable incidents	37265	Dual-chamber implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A05	Mechanical Problem		
									A0709	Device Sensing Problem		
									A070909	Oversensing		
									A0722	Impedance Problem		
									A072201	High Impedance		
									A072202	Low Impedance		
							F26	No Health Consequences or Impact	A05	Mechanical Problem		

									A0709	Device Sensing Problem		
									A070909	Oversensing		
									A0722	Impedance Problem		
									A072201	High Impedance		
									A072202	Low Impedance		
16117	2024	All other reportable incidents	17882	Physiologic-monitoring defibrillation system	CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A0719	Unexpected Shutdown		
							F27	Problem Identified During Non-Clinical Procedure	A0719	Unexpected Shutdown		
16118	2024	All other reportable incidents	17882	Physiologic-monitoring defibrillation system	CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A1603	Failure of Device to Self-Test		
							F27	Problem Identified During Non-Clinical Procedure	A1603	Failure of Device to Self-Test		
16119	2024	All other reportable incidents	37265	Dual-chamber implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A0709	Device Sensing Problem		
									A070909	Oversensing		
									A0712	Pacing Problem		
									A090801	Signal Artifact/Noise		
							F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A0709	Device Sensing Problem		
									A070909	Oversensing		
									A0712	Pacing Problem		
									A090801	Signal Artifact/Noise		
							F12	Serious Injury/ Illness/ Impairment	A0709	Device Sensing Problem		
									A070909	Oversensing		
									A0712	Pacing Problem		
									A090801	Signal Artifact/Noise		
							F2307	Reprogramming of Device	A0709	Device Sensing Problem		
									A070909	Oversensing		
									A0712	Pacing Problem		
									A090801	Signal Artifact/Noise		
16120	2024	All other reportable incidents	47270	Cardiac resynchronization therapy implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E210401	Shock from Patient Lead(s)	A070909	Oversensing	16/12/2024	16/12/2024
									A071302	Inappropriate/Inadequate Shock/Stimulation	16/12/2024	16/12/2024
							F08	Hospitalization or Prolonged Hospitalization	A070909	Oversensing	16/12/2024	16/12/2024
									A071302	Inappropriate/Inadequate Shock/Stimulation	16/12/2024	16/12/2024
							F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A070909	Oversensing	16/12/2024	16/12/2024
									A071302	Inappropriate/Inadequate Shock/Stimulation	16/12/2024	16/12/2024
							F1905	Device Revision or Replacement	A070909	Oversensing	16/12/2024	16/12/2024
									A071302	Inappropriate/Inadequate Shock/Stimulation	16/12/2024	16/12/2024
							F23	Unexpected Medical Intervention	A070909	Oversensing	16/12/2024	16/12/2024
									A071302	Inappropriate/Inadequate Shock/Stimulation	16/12/2024	16/12/2024
16121	2024	All other reportable incidents	47263	Cardiac resynchronization therapy implantable pacemaker	CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A05	Mechanical Problem		
									A072201	High Impedance		
							F26	No Health Consequences or Impact	A05	Mechanical Problem		
									A072201	High Impedance		
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A05	Mechanical Problem		
									A072201	High Impedance		
							F26	No Health Consequences or Impact	A05	Mechanical Problem		
									A072201	High Impedance		
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A05	Mechanical Problem		
									A072201	High Impedance		
							F26	No Health Consequences or Impact	A05	Mechanical Problem		
									A072201	High Impedance		
16122	2024		47265	Dual-chamber implantable pacemaker, rate-responsive	CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	20/09/2024	20/09/2024
							F1901	Additional Surgery	A24	No Apparent Device Problem	20/09/2024	20/09/2024
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	20/09/2024	20/09/2024
							F1901	Additional Surgery	A24	No Apparent Device Problem	20/09/2024	20/09/2024
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	20/09/2024	20/09/2024
							F1901	Additional Surgery	A24	No Apparent Device Problem	20/09/2024	20/09/2024
16123	2024	All other reportable incidents	47265	Dual-chamber implantable pacemaker, rate-responsive	CT1685	Cardiac pacemakers	E0623	Unspecified Heart Problem	A070504	Premature Discharge of Battery	14/11/2024	14/11/2024
									A0908	Incorrect, Inadequate or Imprecise Result or Readings	14/11/2024	14/11/2024
							E2321	Low Blood Pressure/ Hypotension	A070504	Premature Discharge of Battery	14/11/2024	14/11/2024
									A0908	Incorrect, Inadequate or Imprecise Result or Readings	14/11/2024	14/11/2024
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	14/11/2024	14/11/2024
									A0908	Incorrect, Inadequate or Imprecise Result or Readings	14/11/2024	14/11/2024
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	14/11/2024	14/11/2024
									A0908	Incorrect, Inadequate or Imprecise Result or Readings	14/11/2024	14/11/2024
					CT257	Implantable cardiac pacemakers	E0623	Unspecified Heart Problem	A070504	Premature Discharge of Battery	14/11/2024	14/11/2024
									A0908	Incorrect, Inadequate or Imprecise Result or Readings	14/11/2024	14/11/2024
							E2321	Low Blood Pressure/ Hypotension	A070504	Premature Discharge of Battery	14/11/2024	14/11/2024
									A0908	Incorrect, Inadequate or Imprecise Result or Readings	14/11/2024	14/11/2024
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	14/11/2024	14/11/2024
									A0908	Incorrect, Inadequate or Imprecise Result or Readings	14/11/2024	14/11/2024
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	14/11/2024	14/11/2024
									A0908	Incorrect, Inadequate or Imprecise Result or Readings	14/11/2024	14/11/2024
					CT2866	Implantable cardiac electrical stimulators	E0623	Unspecified Heart Problem	A070504	Premature Discharge of Battery	14/11/2024	14/11/2024
									A0908	Incorrect, Inadequate or Imprecise Result or Readings	14/11/2024	14/11/2024
							E2321	Low Blood Pressure/ Hypotension	A070504	Premature Discharge of Battery	14/11/2024	14/11/2024
									A0908	Incorrect, Inadequate or Imprecise Result or Readings	14/11/2024	14/11/2024
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	14/11/2024	14/11/2024
									A0908	Incorrect, Inadequate or Imprecise Result or Readings	14/11/2024	14/11/2024
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	14/11/2024	14/11/2024
									A0908	Incorrect, Inadequate or Imprecise Result or Readings	14/11/2024	14/11/2024
16124	2024	All other reportable incidents	47265	Dual-chamber implantable pacemaker, rate-responsive	CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		

							F26	No Health Consequences or Impact	A0705	Battery Problem			
									A070504	Premature Discharge of Battery			
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem			
									A070504	Premature Discharge of Battery			
							F26	No Health Consequences or Impact	A0705	Battery Problem			
									A070504	Premature Discharge of Battery			
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem			
									A070504	Premature Discharge of Battery			
							F26	No Health Consequences or Impact	A0705	Battery Problem			
									A070504	Premature Discharge of Battery			
16125	2024	All other reportable incidents	47265	Dual-chamber implantable pacemaker, rate-responsive	CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A090805	Incorrect Measurement	07/11/2024	07/11/2024
							F08	Hospitalization or Prolonged Hospitalization	A090805	Incorrect Measurement	07/11/2024	07/11/2024
							F12	Serious Injury/ Illness/ Impairment	A090805	Incorrect Measurement	07/11/2024	07/11/2024
							F1905	Device Revision or Replacement	A090805	Incorrect Measurement	07/11/2024	07/11/2024
							F23	Unexpected Medical Intervention	A090805	Incorrect Measurement	07/11/2024	07/11/2024
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A090805	Incorrect Measurement	07/11/2024	07/11/2024
							F08	Hospitalization or Prolonged Hospitalization	A090805	Incorrect Measurement	07/11/2024	07/11/2024
							F12	Serious Injury/ Illness/ Impairment	A090805	Incorrect Measurement	07/11/2024	07/11/2024
							F1905	Device Revision or Replacement	A090805	Incorrect Measurement	07/11/2024	07/11/2024
							F23	Unexpected Medical Intervention	A090805	Incorrect Measurement	07/11/2024	07/11/2024
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A090805	Incorrect Measurement	07/11/2024	07/11/2024
							F08	Hospitalization or Prolonged Hospitalization	A090805	Incorrect Measurement	07/11/2024	07/11/2024
							F12	Serious Injury/ Illness/ Impairment	A090805	Incorrect Measurement	07/11/2024	07/11/2024
							F1905	Device Revision or Replacement	A090805	Incorrect Measurement	07/11/2024	07/11/2024
							F23	Unexpected Medical Intervention	A090805	Incorrect Measurement	07/11/2024	07/11/2024
16126	2024		47265	Dual-chamber implantable pacemaker, rate-responsive	CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	28/08/2024	28/08/2024
							F1901	Additional Surgery	A24	No Apparent Device Problem	28/08/2024	28/08/2024
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	28/08/2024	28/08/2024
							F1901	Additional Surgery	A24	No Apparent Device Problem	28/08/2024	28/08/2024
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	28/08/2024	28/08/2024
							F1901	Additional Surgery	A24	No Apparent Device Problem	28/08/2024	28/08/2024
16127	2024	All other reportable incidents	60789	Intracardiac pacemaker	CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A05	Mechanical Problem	24/10/2024	24/10/2024
									A051201	Device Dislodged or Dislocated	24/10/2024	24/10/2024
									A1502	Positioning Problem	24/10/2024	24/10/2024
							F26	No Health Consequences or Impact	A05	Mechanical Problem	24/10/2024	24/10/2024
									A051201	Device Dislodged or Dislocated	24/10/2024	24/10/2024
									A1502	Positioning Problem	24/10/2024	24/10/2024
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A05	Mechanical Problem	24/10/2024	24/10/2024
									A051201	Device Dislodged or Dislocated	24/10/2024	24/10/2024
									A1502	Positioning Problem	24/10/2024	24/10/2024
							F26	No Health Consequences or Impact	A05	Mechanical Problem	24/10/2024	24/10/2024
									A051201	Device Dislodged or Dislocated	24/10/2024	24/10/2024
									A1502	Positioning Problem	24/10/2024	24/10/2024
					CT2818	Catheterization/Cannulation/Fluid path devices	E2403	No Clinical Signs, Symptoms or Conditions	A05	Mechanical Problem	24/10/2024	24/10/2024
									A051201	Device Dislodged or Dislocated	24/10/2024	24/10/2024
									A1502	Positioning Problem	24/10/2024	24/10/2024
							F26	No Health Consequences or Impact	A05	Mechanical Problem	24/10/2024	24/10/2024
									A051201	Device Dislodged or Dislocated	24/10/2024	24/10/2024
									A1502	Positioning Problem	24/10/2024	24/10/2024
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A05	Mechanical Problem	24/10/2024	24/10/2024
									A051201	Device Dislodged or Dislocated	24/10/2024	24/10/2024
									A1502	Positioning Problem	24/10/2024	24/10/2024
							F26	No Health Consequences or Impact	A05	Mechanical Problem	24/10/2024	24/10/2024
									A051201	Device Dislodged or Dislocated	24/10/2024	24/10/2024
									A1502	Positioning Problem	24/10/2024	24/10/2024
16128	2024	All other reportable incidents	47263	Cardiac resynchronization therapy implantable pacemaker	CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A0908	Incorrect, Inadequate or Imprecise Result or Readings	24/10/2024	24/10/2024
							F08	Hospitalization or Prolonged Hospitalization	A0908	Incorrect, Inadequate or Imprecise Result or Readings	24/10/2024	24/10/2024
							F1905	Device Revision or Replacement	A0908	Incorrect, Inadequate or Imprecise Result or Readings	24/10/2024	24/10/2024
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A0908	Incorrect, Inadequate or Imprecise Result or Readings	24/10/2024	24/10/2024
							F08	Hospitalization or Prolonged Hospitalization	A0908	Incorrect, Inadequate or Imprecise Result or Readings	24/10/2024	24/10/2024
							F1905	Device Revision or Replacement	A0908	Incorrect, Inadequate or Imprecise Result or Readings	24/10/2024	24/10/2024
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A0908	Incorrect, Inadequate or Imprecise Result or Readings	24/10/2024	24/10/2024
							F08	Hospitalization or Prolonged Hospitalization	A0908	Incorrect, Inadequate or Imprecise Result or Readings	24/10/2024	24/10/2024
							F1905	Device Revision or Replacement	A0908	Incorrect, Inadequate or Imprecise Result or Readings	24/10/2024	24/10/2024
16129	2024	All other reportable incidents	47263	Cardiac resynchronization therapy implantable pacemaker	CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem	21/08/2024	21/08/2024
									A070504	Premature Discharge of Battery	21/08/2024	21/08/2024
							F08	Hospitalization or Prolonged Hospitalization	A0705	Battery Problem	21/08/2024	21/08/2024
									A070504	Premature Discharge of Battery	21/08/2024	21/08/2024
							F1905	Device Revision or Replacement	A0705	Battery Problem	21/08/2024	21/08/2024
									A070504	Premature Discharge of Battery	21/08/2024	21/08/2024
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem	21/08/2024	21/08/2024
									A070504	Premature Discharge of Battery	21/08/2024	21/08/2024
							F08	Hospitalization or Prolonged Hospitalization	A0705	Battery Problem	21/08/2024	21/08/2024
									A070504	Premature Discharge of Battery	21/08/2024	21/08/2024
							F1905	Device Revision or Replacement	A0705	Battery Problem	21/08/2024	21/08/2024
									A070504	Premature Discharge of Battery	21/08/2024	21/08/2024
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem	21/08/2024	21/08/2024
									A070504	Premature Discharge of Battery	21/08/2024	21/08/2024
							F08	Hospitalization or Prolonged Hospitalization	A0705	Battery Problem	21/08/2024	21/08/2024
									A070504	Premature Discharge of Battery	21/08/2024	21/08/2024
							F1905	Device Revision or Replacement	A0705	Battery Problem	21/08/2024	21/08/2024

									A070504	Premature Discharge of Battery	21/08/2024	21/08/2024
16130	2024	All other reportable incidents	47265	Dual-chamber implantable pacemaker, rate-responsive	CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
16131	2024	All other reportable incidents	47270	Cardiac resynchronization therapy implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	22/01/2025	22/01/2025
									A0908	Incorrect, Inadequate or Imprecise Result or Readings	22/01/2025	22/01/2025
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	22/01/2025	22/01/2025
									A0908	Incorrect, Inadequate or Imprecise Result or Readings	22/01/2025	22/01/2025
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	22/01/2025	22/01/2025
									A0908	Incorrect, Inadequate or Imprecise Result or Readings	22/01/2025	22/01/2025
16132	2024	All other reportable incidents	47270	Cardiac resynchronization therapy implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	04/06/2025	04/06/2025
									A0908	Incorrect, Inadequate or Imprecise Result or Readings	04/06/2025	04/06/2025
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	04/06/2025	04/06/2025
									A0908	Incorrect, Inadequate or Imprecise Result or Readings	04/06/2025	04/06/2025
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	04/06/2025	04/06/2025
									A0908	Incorrect, Inadequate or Imprecise Result or Readings	04/06/2025	04/06/2025
16133	2024	All other reportable incidents	47265	Dual-chamber implantable pacemaker, rate-responsive	CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A13	Communication or Transmission Problem		
							F26	No Health Consequences or Impact	A13	Communication or Transmission Problem		
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A13	Communication or Transmission Problem		
							F26	No Health Consequences or Impact	A13	Communication or Transmission Problem		
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A13	Communication or Transmission Problem		
							F26	No Health Consequences or Impact	A13	Communication or Transmission Problem		
16134	2024	All other reportable incidents	47265	Dual-chamber implantable pacemaker, rate-responsive	CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A0908	Incorrect, Inadequate or Imprecise Result or Readings	22/11/2024	22/11/2024
							F08	Hospitalization or Prolonged Hospitalization	A0908	Incorrect, Inadequate or Imprecise Result or Readings	22/11/2024	22/11/2024
							F1905	Device Revision or Replacement	A0908	Incorrect, Inadequate or Imprecise Result or Readings	22/11/2024	22/11/2024
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A0908	Incorrect, Inadequate or Imprecise Result or Readings	22/11/2024	22/11/2024
							F08	Hospitalization or Prolonged Hospitalization	A0908	Incorrect, Inadequate or Imprecise Result or Readings	22/11/2024	22/11/2024
							F1905	Device Revision or Replacement	A0908	Incorrect, Inadequate or Imprecise Result or Readings	22/11/2024	22/11/2024
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A0908	Incorrect, Inadequate or Imprecise Result or Readings	22/11/2024	22/11/2024
							F08	Hospitalization or Prolonged Hospitalization	A0908	Incorrect, Inadequate or Imprecise Result or Readings	22/11/2024	22/11/2024
							F1905	Device Revision or Replacement	A0908	Incorrect, Inadequate or Imprecise Result or Readings	22/11/2024	22/11/2024
16135	2024	All other reportable incidents	37265	Dual-chamber implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E210401	Shock from Patient Lead(s)	A0712	Pacing Problem		
									A071302	Inappropriate/Inadequate Shock/Stimulation		
									A1301	Failure to Read Input Signal		
							F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A0712	Pacing Problem		
									A071302	Inappropriate/Inadequate Shock/Stimulation		
									A1301	Failure to Read Input Signal		
16136	2024	Unanticipated serious deterioration in state of health	47265	Dual-chamber implantable pacemaker, rate-responsive	CT1685	Cardiac pacemakers	E011903	Syncope/Fainting	A0712	Pacing Problem	10/08/2024	10/08/2024
							E060101	Asystole	A0712	Pacing Problem	10/08/2024	10/08/2024
							E0602	Cardiac Arrest	A0712	Pacing Problem	10/08/2024	10/08/2024
							F08	Hospitalization or Prolonged Hospitalization	A0712	Pacing Problem	10/08/2024	10/08/2024
							F1203	Life Threatening Illness or Injury	A0712	Pacing Problem	10/08/2024	10/08/2024
							F1905	Device Revision or Replacement	A0712	Pacing Problem	10/08/2024	10/08/2024
					CT257	Implantable cardiac pacemakers	E011903	Syncope/Fainting	A0712	Pacing Problem	10/08/2024	10/08/2024
							E060101	Asystole	A0712	Pacing Problem	10/08/2024	10/08/2024
							E0602	Cardiac Arrest	A0712	Pacing Problem	10/08/2024	10/08/2024
							F08	Hospitalization or Prolonged Hospitalization	A0712	Pacing Problem	10/08/2024	10/08/2024
							F1203	Life Threatening Illness or Injury	A0712	Pacing Problem	10/08/2024	10/08/2024
							F1905	Device Revision or Replacement	A0712	Pacing Problem	10/08/2024	10/08/2024
					CT2866	Implantable cardiac electrical stimulators	E011903	Syncope/Fainting	A0712	Pacing Problem	10/08/2024	10/08/2024
							E060101	Asystole	A0712	Pacing Problem	10/08/2024	10/08/2024
							E0602	Cardiac Arrest	A0712	Pacing Problem	10/08/2024	10/08/2024
							F08	Hospitalization or Prolonged Hospitalization	A0712	Pacing Problem	10/08/2024	10/08/2024
							F1203	Life Threatening Illness or Injury	A0712	Pacing Problem	10/08/2024	10/08/2024
							F1905	Device Revision or Replacement	A0712	Pacing Problem	10/08/2024	10/08/2024
16137	2024	All other reportable incidents	17882	Physiologic-monitoring defibrillation system	CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A1603	Failure of Device to Self-Test		
							F27	Problem Identified During Non-Clinical Procedure	A1603	Failure of Device to Self-Test		
16138	2024	All other reportable incidents	47270	Cardiac resynchronization therapy implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A072201	High Impedance		
							F26	No Health Consequences or Impact	A072201	High Impedance		
16139	2024	All other reportable incidents	37265	Dual-chamber implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A070909	Oversensing		
							F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A070909	Oversensing		
16140	2024	All other reportable incidents	47265	Dual-chamber implantable pacemaker, rate-responsive	CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A0712	Pacing Problem		
									A072201	High Impedance		
									A1301	Failure to Read Input Signal		
							F26	No Health Consequences or Impact	A0712	Pacing Problem		
									A072201	High Impedance		
									A1301	Failure to Read Input Signal		
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A0712	Pacing Problem		
									A072201	High Impedance		
									A1301	Failure to Read Input Signal		
							F26	No Health Consequences or Impact	A0712	Pacing Problem		
									A072201	High Impedance		
									A1301	Failure to Read Input Signal		

					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A0712	Pacing Problem		
									A072201	High Impedance		
									A1301	Failure to Read Input Signal		
							F26	No Health Consequences or Impact	A0712	Pacing Problem		
									A072201	High Impedance		
									A1301	Failure to Read Input Signal		
16141	2024	All other reportable incidents	17882	Physiologic-monitoring defibrillation system	CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A0719	Unexpected Shutdown		
							F26	No Health Consequences or Impact	A0719	Unexpected Shutdown		
16142	2024	All other reportable incidents	37265	Dual-chamber implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A0908	Incorrect, Inadequate or Imprecise Result or Readings		
							F26	No Health Consequences or Impact	A0908	Incorrect, Inadequate or Imprecise Result or Readings		
16143	2024	All other reportable incidents	17882	Physiologic-monitoring defibrillation system	CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A07	Electrical /Electronic Property Problem		
							F26	No Health Consequences or Impact	A07	Electrical /Electronic Property Problem		
16144	2024	All other reportable incidents	35852	Single-chamber implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E210401	Shock from Patient Lead(s)	A070102	High Capture Threshold	08/11/2024	08/11/2024
									A070909	Oversensing	08/11/2024	08/11/2024
									A071302	Inappropriate/Inadequate Shock/Stimulation	08/11/2024	08/11/2024
									A0908	Incorrect, Inadequate or Imprecise Result or Readings	08/11/2024	08/11/2024
									A090801	Signal Artifact/Noise	08/11/2024	08/11/2024
							F08	Hospitalization or Prolonged Hospitalization	A070102	High Capture Threshold	08/11/2024	08/11/2024
									A070909	Oversensing	08/11/2024	08/11/2024
									A071302	Inappropriate/Inadequate Shock/Stimulation	08/11/2024	08/11/2024
									A0908	Incorrect, Inadequate or Imprecise Result or Readings	08/11/2024	08/11/2024
									A090801	Signal Artifact/Noise	08/11/2024	08/11/2024
							F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A070102	High Capture Threshold	08/11/2024	08/11/2024
									A070909	Oversensing	08/11/2024	08/11/2024
									A071302	Inappropriate/Inadequate Shock/Stimulation	08/11/2024	08/11/2024
									A0908	Incorrect, Inadequate or Imprecise Result or Readings	08/11/2024	08/11/2024
									A090801	Signal Artifact/Noise	08/11/2024	08/11/2024
							F1905	Device Revision or Replacement	A070102	High Capture Threshold	08/11/2024	08/11/2024
									A070909	Oversensing	08/11/2024	08/11/2024
									A071302	Inappropriate/Inadequate Shock/Stimulation	08/11/2024	08/11/2024
									A0908	Incorrect, Inadequate or Imprecise Result or Readings	08/11/2024	08/11/2024
									A090801	Signal Artifact/Noise	08/11/2024	08/11/2024
							F2203	Imaging Required	A070102	High Capture Threshold	08/11/2024	08/11/2024
									A070909	Oversensing	08/11/2024	08/11/2024
									A071302	Inappropriate/Inadequate Shock/Stimulation	08/11/2024	08/11/2024
									A0908	Incorrect, Inadequate or Imprecise Result or Readings	08/11/2024	08/11/2024
									A090801	Signal Artifact/Noise	08/11/2024	08/11/2024
16145	2024		47265	Dual-chamber implantable pacemaker, rate-responsive	CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	17/09/2024	17/09/2024
							F1901	Additional Surgery	A24	No Apparent Device Problem	17/09/2024	17/09/2024
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	17/09/2024	17/09/2024
							F1901	Additional Surgery	A24	No Apparent Device Problem	17/09/2024	17/09/2024
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	17/09/2024	17/09/2024
							F1901	Additional Surgery	A24	No Apparent Device Problem	17/09/2024	17/09/2024
16146	2024	All other reportable incidents	47270	Cardiac resynchronization therapy implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	07/01/2025	07/01/2025
									A0908	Incorrect, Inadequate or Imprecise Result or Readings	07/01/2025	07/01/2025
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	07/01/2025	07/01/2025
									A0908	Incorrect, Inadequate or Imprecise Result or Readings	07/01/2025	07/01/2025
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	07/01/2025	07/01/2025
									A0908	Incorrect, Inadequate or Imprecise Result or Readings	07/01/2025	07/01/2025
16147	2024	All other reportable incidents	47270	Cardiac resynchronization therapy implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E210401	Shock from Patient Lead(s)	A070909	Oversensing		
									A071302	Inappropriate/Inadequate Shock/Stimulation		
									A090801	Signal Artifact/Noise		
									A1301	Failure to Read Input Signal		
							F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A070909	Oversensing		
									A071302	Inappropriate/Inadequate Shock/Stimulation		
									A090801	Signal Artifact/Noise		
									A1301	Failure to Read Input Signal		
							F2203	Imaging Required	A070909	Oversensing		
									A071302	Inappropriate/Inadequate Shock/Stimulation		
									A090801	Signal Artifact/Noise		
									A1301	Failure to Read Input Signal		
16148	2024	Death	47265	Dual-chamber implantable pacemaker, rate-responsive	CT1685	Cardiac pacemakers	E2401	Insufficient Information	A24	No Apparent Device Problem		
							F02	Death	A24	No Apparent Device Problem		
					CT257	Implantable cardiac pacemakers	E2401	Insufficient Information	A24	No Apparent Device Problem		
							F02	Death	A24	No Apparent Device Problem		
					CT2866	Implantable cardiac electrical stimulators	E2401	Insufficient Information	A24	No Apparent Device Problem		
							F02	Death	A24	No Apparent Device Problem		
16149	2024	All other reportable incidents			CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070101	Failure to Capture		
									A070909	Oversensing		
									A070910	Undersensing		
									A090801	Signal Artifact/Noise		
							F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A070101	Failure to Capture		
									A070909	Oversensing		
									A070910	Undersensing		
									A090801	Signal Artifact/Noise		
16150	2024		47265	Dual-chamber implantable pacemaker, rate-responsive	CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	06/08/2024	06/08/2024
							F1901	Additional Surgery	A24	No Apparent Device Problem	06/08/2024	06/08/2024
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	06/08/2024	06/08/2024
							F1901	Additional Surgery	A24	No Apparent Device Problem	06/08/2024	06/08/2024
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	06/08/2024	06/08/2024

							F1901	Additional Surgery	A24	No Apparent Device Problem	06/08/2024	06/08/2024
16151	2024	All other reportable incidents	37265	Dual-chamber implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A072201	High Impedance		
							F26	No Health Consequences or Impact	A072201	High Impedance		
16152	2024	All other reportable incidents	47270	Cardiac resynchronization therapy implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A070909	Oversensing	11/10/2024	11/10/2024
							F08	Hospitalization or Prolonged Hospitalization	A070909	Oversensing	11/10/2024	11/10/2024
							F1905	Device Revision or Replacement	A070909	Oversensing	11/10/2024	11/10/2024
16153	2024	All other reportable incidents	47265	Dual-chamber implantable pacemaker, rate-responsive	CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
16154	2024	All other reportable incidents	17882	Physiologic-monitoring defibrillation system	CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A0904	Energy Output Problem		
							F27	Problem Identified During Non-Clinical Procedure	A0904	Energy Output Problem		
16155	2024	All other reportable incidents	17882	Physiologic-monitoring defibrillation system	CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070912	Incorrect Interpretation of Signal		
							F26	No Health Consequences or Impact	A070912	Incorrect Interpretation of Signal		
16156	2024	All other reportable incidents	47263	Cardiac resynchronization therapy implantable pacemaker	CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A0908	Incorrect, Inadequate or Imprecise Result or Readings	29/10/2024	29/10/2024
							F08	Hospitalization or Prolonged Hospitalization	A0908	Incorrect, Inadequate or Imprecise Result or Readings	29/10/2024	29/10/2024
							F1905	Device Revision or Replacement	A0908	Incorrect, Inadequate or Imprecise Result or Readings	29/10/2024	29/10/2024
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A0908	Incorrect, Inadequate or Imprecise Result or Readings	29/10/2024	29/10/2024
							F08	Hospitalization or Prolonged Hospitalization	A0908	Incorrect, Inadequate or Imprecise Result or Readings	29/10/2024	29/10/2024
							F1905	Device Revision or Replacement	A0908	Incorrect, Inadequate or Imprecise Result or Readings	29/10/2024	29/10/2024
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A0908	Incorrect, Inadequate or Imprecise Result or Readings	29/10/2024	29/10/2024
							F08	Hospitalization or Prolonged Hospitalization	A0908	Incorrect, Inadequate or Imprecise Result or Readings	29/10/2024	29/10/2024
							F1905	Device Revision or Replacement	A0908	Incorrect, Inadequate or Imprecise Result or Readings	29/10/2024	29/10/2024
16157	2024	All other reportable incidents	47263	Cardiac resynchronization therapy implantable pacemaker	CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A0908	Incorrect, Inadequate or Imprecise Result or Readings		
							F26	No Health Consequences or Impact	A0908	Incorrect, Inadequate or Imprecise Result or Readings		
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A0908	Incorrect, Inadequate or Imprecise Result or Readings		
							F26	No Health Consequences or Impact	A0908	Incorrect, Inadequate or Imprecise Result or Readings		
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A0908	Incorrect, Inadequate or Imprecise Result or Readings		
							F26	No Health Consequences or Impact	A0908	Incorrect, Inadequate or Imprecise Result or Readings		
16158	2024	All other reportable incidents	47265	Dual-chamber implantable pacemaker, rate-responsive	CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A070910	Undersensing		
							F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A070910	Undersensing		
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A070910	Undersensing		
							F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A070910	Undersensing		
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A070910	Undersensing		
							F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A070910	Undersensing		
16159	2024	All other reportable incidents	47265	Dual-chamber implantable pacemaker, rate-responsive	CT1685	Cardiac pacemakers	E0623	Unspecified Heart Problem	A0908	Incorrect, Inadequate or Imprecise Result or Readings	08/05/2024	08/05/2024
							F08	Hospitalization or Prolonged Hospitalization	A0908	Incorrect, Inadequate or Imprecise Result or Readings	08/05/2024	08/05/2024
							F1905	Device Revision or Replacement	A0908	Incorrect, Inadequate or Imprecise Result or Readings	08/05/2024	08/05/2024
					CT257	Implantable cardiac pacemakers	E0623	Unspecified Heart Problem	A0908	Incorrect, Inadequate or Imprecise Result or Readings	08/05/2024	08/05/2024
							F08	Hospitalization or Prolonged Hospitalization	A0908	Incorrect, Inadequate or Imprecise Result or Readings	08/05/2024	08/05/2024
							F1905	Device Revision or Replacement	A0908	Incorrect, Inadequate or Imprecise Result or Readings	08/05/2024	08/05/2024
					CT2866	Implantable cardiac electrical stimulators	E0623	Unspecified Heart Problem	A0908	Incorrect, Inadequate or Imprecise Result or Readings	08/05/2024	08/05/2024
							F08	Hospitalization or Prolonged Hospitalization	A0908	Incorrect, Inadequate or Imprecise Result or Readings	08/05/2024	08/05/2024
							F1905	Device Revision or Replacement	A0908	Incorrect, Inadequate or Imprecise Result or Readings	08/05/2024	08/05/2024
16160	2024	All other reportable incidents	37265	Dual-chamber implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A0712	Pacing Problem		
							F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A0712	Pacing Problem		
16161	2024	All other reportable incidents	47270	Cardiac resynchronization therapy implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E0601	Arrhythmia	A070909	Oversensing		
									A071201	Failure to Convert Rhythm		
							F12	Serious Injury/ Illness/ Impairment	A070909	Oversensing		
									A071201	Failure to Convert Rhythm		
16162	2024	All other reportable incidents	47205	Cardiac pulse generator programmer	CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A09	Output Problem		
									A1104	Computer Operating System Problem		
									A13	Communication or Transmission Problem		
							F26	No Health Consequences or Impact	A09	Output Problem		
									A1104	Computer Operating System Problem		
									A13	Communication or Transmission Problem		
16163	2024		47263	Cardiac resynchronization therapy implantable pacemaker	CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A071205	Pacemaker Found in Back-Up Mode	29/10/2024	
							F1905	Device Revision or Replacement	A071205	Pacemaker Found in Back-Up Mode	29/10/2024	
							F2307	Reprogramming of Device	A071205	Pacemaker Found in Back-Up Mode	29/10/2024	
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A071205	Pacemaker Found in Back-Up Mode	29/10/2024	
							F1905	Device Revision or Replacement	A071205	Pacemaker Found in Back-Up Mode	29/10/2024	
							F2307	Reprogramming of Device	A071205	Pacemaker Found in Back-Up Mode	29/10/2024	
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A071205	Pacemaker Found in Back-Up Mode	29/10/2024	
							F1905	Device Revision or Replacement	A071205	Pacemaker Found in Back-Up Mode	29/10/2024	
							F2307	Reprogramming of Device	A071205	Pacemaker Found in Back-Up Mode	29/10/2024	
16164	2024	All other reportable incidents	47265	Dual-chamber implantable pacemaker, rate-responsive	CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	28/10/2024	28/10/2024
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	28/10/2024	28/10/2024

							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	28/10/2024	28/10/2024
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	28/10/2024	28/10/2024
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	28/10/2024	28/10/2024
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	28/10/2024	28/10/2024
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	28/10/2024	28/10/2024
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	28/10/2024	28/10/2024
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	28/10/2024	28/10/2024
16165	2024	All other reportable incidents	47265	Dual-chamber implantable pacemaker, rate-responsive	CT1685	Cardiac pacemakers	E2401	Insufficient Information	A12	Connection Problem	20/08/2024	
							F1903	Device Explantation	A12	Connection Problem	20/08/2024	
					CT257	Implantable cardiac pacemakers	E2401	Insufficient Information	A12	Connection Problem	20/08/2024	
							F1903	Device Explantation	A12	Connection Problem	20/08/2024	
					CT2866	Implantable cardiac electrical stimulators	E2401	Insufficient Information	A12	Connection Problem	20/08/2024	
							F1903	Device Explantation	A12	Connection Problem	20/08/2024	
16166	2024	All other reportable incidents			CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
16167	2024	All other reportable incidents			CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070910	Undersensing		
									A0712	Pacing Problem		
							F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A070910	Undersensing		
									A0712	Pacing Problem		
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A070910	Undersensing		
									A0712	Pacing Problem		
							F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A070910	Undersensing		
									A0712	Pacing Problem		
16168	2024	All other reportable incidents	35852	Single-chamber implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E210401	Shock from Patient Lead(s)	A070909	Oversensing		
									A071302	Inappropriate/Inadequate Shock/Stimulation		
							F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A070909	Oversensing		
									A071302	Inappropriate/Inadequate Shock/Stimulation		
16169	2024	All other reportable incidents	47265	Dual-chamber implantable pacemaker, rate-responsive	CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A0908	Incorrect, Inadequate or Imprecise Result or Readings	04/11/2024	04/11/2024
							F26	No Health Consequences or Impact	A0908	Incorrect, Inadequate or Imprecise Result or Readings	04/11/2024	04/11/2024
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A0908	Incorrect, Inadequate or Imprecise Result or Readings	04/11/2024	04/11/2024
							F26	No Health Consequences or Impact	A0908	Incorrect, Inadequate or Imprecise Result or Readings	04/11/2024	04/11/2024
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A0908	Incorrect, Inadequate or Imprecise Result or Readings	04/11/2024	04/11/2024
							F26	No Health Consequences or Impact	A0908	Incorrect, Inadequate or Imprecise Result or Readings	04/11/2024	04/11/2024
16170	2024	All other reportable incidents	37265	Dual-chamber implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery		
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery		
16171	2024		47265	Dual-chamber implantable pacemaker, rate-responsive	CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	17/09/2024	17/09/2024
							F1901	Additional Surgery	A24	No Apparent Device Problem	17/09/2024	17/09/2024
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	17/09/2024	17/09/2024
							F1901	Additional Surgery	A24	No Apparent Device Problem	17/09/2024	17/09/2024
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	17/09/2024	17/09/2024
							F1901	Additional Surgery	A24	No Apparent Device Problem	17/09/2024	17/09/2024
16172	2024		47265	Dual-chamber implantable pacemaker, rate-responsive	CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	17/09/2024	17/09/2024
							F1901	Additional Surgery	A24	No Apparent Device Problem	17/09/2024	17/09/2024
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	17/09/2024	17/09/2024
							F1901	Additional Surgery	A24	No Apparent Device Problem	17/09/2024	17/09/2024
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	17/09/2024	17/09/2024
							F1901	Additional Surgery	A24	No Apparent Device Problem	17/09/2024	17/09/2024
16173	2024	All other reportable incidents	47265	Dual-chamber implantable pacemaker, rate-responsive	CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	07/02/2024	07/02/2024
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	07/02/2024	07/02/2024
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	07/02/2024	07/02/2024
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	07/02/2024	07/02/2024
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	07/02/2024	07/02/2024
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	07/02/2024	07/02/2024
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	07/02/2024	07/02/2024
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	07/02/2024	07/02/2024
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	07/02/2024	07/02/2024
16174	2024	All other reportable incidents	47270	Cardiac resynchronization therapy implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A070909	Oversensing		
							F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A070909	Oversensing		
16175	2024	All other reportable incidents	47270	Cardiac resynchronization therapy implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	23/01/2025	23/01/2025
									A0908	Incorrect, Inadequate or Imprecise Result or Readings	23/01/2025	23/01/2025
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	23/01/2025	23/01/2025
									A0908	Incorrect, Inadequate or Imprecise Result or Readings	23/01/2025	23/01/2025
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	23/01/2025	23/01/2025
									A0908	Incorrect, Inadequate or Imprecise Result or Readings	23/01/2025	23/01/2025
16176	2024	All other reportable incidents	47265	Dual-chamber implantable pacemaker, rate-responsive	CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery		
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery		
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery		
16177	2024	All other reportable incidents	47263	Cardiac resynchronization therapy implantable pacemaker	CT1685	Cardiac pacemakers	E060101	Asystole	A070101	Failure to Capture		
									A070909	Oversensing		
									A0712	Pacing Problem		
									A072201	High Impedance		

									A090801	Signal Artifact/Noise		
							F1001	Absence of Treatment	A070101	Failure to Capture		
									A070909	Oversensing		
									A0712	Pacing Problem		
									A072201	High Impedance		
									A090801	Signal Artifact/Noise		
					CT257	Implantable cardiac pacemakers	E060101	Asystole	A070101	Failure to Capture		
									A070909	Oversensing		
									A0712	Pacing Problem		
									A072201	High Impedance		
									A090801	Signal Artifact/Noise		
							F1001	Absence of Treatment	A070101	Failure to Capture		
									A070909	Oversensing		
									A0712	Pacing Problem		
									A072201	High Impedance		
									A090801	Signal Artifact/Noise		
					CT2866	Implantable cardiac electrical stimulators	E060101	Asystole	A070101	Failure to Capture		
									A070909	Oversensing		
									A0712	Pacing Problem		
									A072201	High Impedance		
									A090801	Signal Artifact/Noise		
							F1001	Absence of Treatment	A070101	Failure to Capture		
									A070909	Oversensing		
									A0712	Pacing Problem		
									A072201	High Impedance		
									A090801	Signal Artifact/Noise		
16178	2024	All other reportable incidents	47265	Dual-chamber implantable pacemaker, rate-responsive	CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A071102	Failure to Interrogate		
							F26	No Health Consequences or Impact	A071102	Failure to Interrogate		
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A071102	Failure to Interrogate		
							F26	No Health Consequences or Impact	A071102	Failure to Interrogate		
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A071102	Failure to Interrogate		
							F26	No Health Consequences or Impact	A071102	Failure to Interrogate		
16179	2024	Unanticipated serious deterioration in state of health	47270	Cardiac resynchronization therapy implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E011903	Syncope/Fainting	A0709	Device Sensing Problem		
									A071301	Failure to Deliver Shock/Stimulation		
							E060109	Tachycardia	A0709	Device Sensing Problem		
									A071301	Failure to Deliver Shock/Stimulation		
							F1203	Life Threatening Illness or Injury	A0709	Device Sensing Problem		
									A071301	Failure to Deliver Shock/Stimulation		
							F2307	Reprogramming of Device	A0709	Device Sensing Problem		
									A071301	Failure to Deliver Shock/Stimulation		
16180	2024		47265	Dual-chamber implantable pacemaker, rate-responsive	CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	17/09/2024	17/09/2024
							F1901	Additional Surgery	A24	No Apparent Device Problem	17/09/2024	17/09/2024
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	17/09/2024	17/09/2024
							F1901	Additional Surgery	A24	No Apparent Device Problem	17/09/2024	17/09/2024
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	17/09/2024	17/09/2024
							F1901	Additional Surgery	A24	No Apparent Device Problem	17/09/2024	17/09/2024
16181	2024	All other reportable incidents	47270	Cardiac resynchronization therapy implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	06/01/2025	06/01/2025
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery	06/01/2025	06/01/2025
16182	2024	All other reportable incidents	47270	Cardiac resynchronization therapy implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A05	Mechanical Problem	01/11/2024	01/11/2024
							F08	Hospitalization or Prolonged Hospitalization	A05	Mechanical Problem	01/11/2024	01/11/2024
							F12	Serious Injury/ Illness/ Impairment	A05	Mechanical Problem	01/11/2024	01/11/2024
							F1905	Device Revision or Replacement	A05	Mechanical Problem	01/11/2024	01/11/2024
16183	2024	All other reportable incidents	17882	Physiologic-monitoring defibrillation system	CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A071301	Failure to Deliver Shock/Stimulation		
							F26	No Health Consequences or Impact	A071301	Failure to Deliver Shock/Stimulation		
16184	2024	All other reportable incidents	17882	Physiologic-monitoring defibrillation system	CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A0719	Unexpected Shutdown		
							F27	Problem Identified During Non-Clinical Procedure	A0719	Unexpected Shutdown		
16185	2024	All other reportable incidents	36106	Implantable pacemaker analyser, intraoperative	CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A0719	Unexpected Shutdown		
							F27	Problem Identified During Non-Clinical Procedure	A0719	Unexpected Shutdown		
16186	2024	All other reportable incidents	47265	Dual-chamber implantable pacemaker, rate-responsive	CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A0203	Defective Device	22/11/2024	22/11/2024
									A0908	Incorrect, Inadequate or Imprecise Result or Readings	22/11/2024	22/11/2024
							F08	Hospitalization or Prolonged Hospitalization	A0203	Defective Device	22/11/2024	22/11/2024
									A0908	Incorrect, Inadequate or Imprecise Result or Readings	22/11/2024	22/11/2024
							F1905	Device Revision or Replacement	A0203	Defective Device	22/11/2024	22/11/2024
									A0908	Incorrect, Inadequate or Imprecise Result or Readings	22/11/2024	22/11/2024
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A0203	Defective Device	22/11/2024	22/11/2024
									A0908	Incorrect, Inadequate or Imprecise Result or Readings	22/11/2024	22/11/2024
							F08	Hospitalization or Prolonged Hospitalization	A0203	Defective Device	22/11/2024	22/11/2024
									A0908	Incorrect, Inadequate or Imprecise Result or Readings	22/11/2024	22/11/2024
							F1905	Device Revision or Replacement	A0203	Defective Device	22/11/2024	22/11/2024
									A0908	Incorrect, Inadequate or Imprecise Result or Readings	22/11/2024	22/11/2024
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A0203	Defective Device	22/11/2024	22/11/2024
									A0908	Incorrect, Inadequate or Imprecise Result or Readings	22/11/2024	22/11/2024
							F08	Hospitalization or Prolonged Hospitalization	A0203	Defective Device	22/11/2024	22/11/2024
									A0908	Incorrect, Inadequate or Imprecise Result or Readings	22/11/2024	22/11/2024
							F1905	Device Revision or Replacement	A0203	Defective Device	22/11/2024	22/11/2024
									A0908	Incorrect, Inadequate or Imprecise Result or Readings	22/11/2024	22/11/2024
16187	2024	All other reportable incidents	47270	Cardiac resynchronization therapy implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
16188	2024	All other reportable incidents	17882	Physiologic-monitoring defibrillation system	CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070803	Failure to Power Up		
							F27	Problem Identified During Non-Clinical Procedure	A070803	Failure to Power Up		
16189	2024	All other reportable incidents	47265	Dual-chamber implantable pacemaker, rate-	CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem		

				responsive					A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
16190	2024	All other reportable incidents	47265	Dual-chamber implantable pacemaker, rate-responsive	CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A070102	High Capture Threshold		
									A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070102	High Capture Threshold		
									A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A070102	High Capture Threshold		
									A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070102	High Capture Threshold		
									A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A070102	High Capture Threshold		
									A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070102	High Capture Threshold		
									A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
16191	2024	All other reportable incidents	37265	Dual-chamber implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A070909	Oversensing		
									A070910	Undersensing		
									A0712	Pacing Problem		
							F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A070909	Oversensing		
									A070910	Undersensing		
									A0712	Pacing Problem		
16192	2024	All other reportable incidents	35852	Single-chamber implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A1605	Reset Problem		
									A160503	Inappropriate or Unexpected Reset		
							F12	Serious Injury/ Illness/ Impairment	A1605	Reset Problem		
									A160503	Inappropriate or Unexpected Reset		
							F2307	Reprogramming of Device	A1605	Reset Problem		
									A160503	Inappropriate or Unexpected Reset		
16193	2024	All other reportable incidents	47263	Cardiac resynchronization therapy implantable pacemaker	CT1685	Cardiac pacemakers	E2103	Device Overstimulation of Tissue	A070504	Premature Discharge of Battery	28/11/2024	28/11/2024
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	28/11/2024	28/11/2024
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	28/11/2024	28/11/2024
					CT257	Implantable cardiac pacemakers	E2103	Device Overstimulation of Tissue	A070504	Premature Discharge of Battery	28/11/2024	28/11/2024
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	28/11/2024	28/11/2024
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	28/11/2024	28/11/2024
					CT2866	Implantable cardiac electrical stimulators	E2103	Device Overstimulation of Tissue	A070504	Premature Discharge of Battery	28/11/2024	28/11/2024
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	28/11/2024	28/11/2024
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	28/11/2024	28/11/2024
16194	2024	All other reportable incidents	47205	Cardiac pulse generator programmer	CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A13	Communication or Transmission Problem		
							F26	No Health Consequences or Impact	A13	Communication or Transmission Problem		
16195	2024	All other reportable incidents	47265	Dual-chamber implantable pacemaker, rate-responsive	CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A0712	Pacing Problem	01/11/2024	01/11/2024
									A0908	Incorrect, Inadequate or Imprecise Result or Readings	01/11/2024	01/11/2024
									A1304	Telemetry Discrepancy	01/11/2024	01/11/2024
							F08	Hospitalization or Prolonged Hospitalization	A0712	Pacing Problem	01/11/2024	01/11/2024
									A0908	Incorrect, Inadequate or Imprecise Result or Readings	01/11/2024	01/11/2024
									A1304	Telemetry Discrepancy	01/11/2024	01/11/2024
							F1905	Device Revision or Replacement	A0712	Pacing Problem	01/11/2024	01/11/2024
									A0908	Incorrect, Inadequate or Imprecise Result or Readings	01/11/2024	01/11/2024
									A1304	Telemetry Discrepancy	01/11/2024	01/11/2024
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A0712	Pacing Problem	01/11/2024	01/11/2024
									A0908	Incorrect, Inadequate or Imprecise Result or Readings	01/11/2024	01/11/2024
									A1304	Telemetry Discrepancy	01/11/2024	01/11/2024
							F08	Hospitalization or Prolonged Hospitalization	A0712	Pacing Problem	01/11/2024	01/11/2024
									A0908	Incorrect, Inadequate or Imprecise Result or Readings	01/11/2024	01/11/2024
									A1304	Telemetry Discrepancy	01/11/2024	01/11/2024
							F1905	Device Revision or Replacement	A0712	Pacing Problem	01/11/2024	01/11/2024
									A0908	Incorrect, Inadequate or Imprecise Result or Readings	01/11/2024	01/11/2024
									A1304	Telemetry Discrepancy	01/11/2024	01/11/2024
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A0712	Pacing Problem	01/11/2024	01/11/2024
									A0908	Incorrect, Inadequate or Imprecise Result or Readings	01/11/2024	01/11/2024
									A1304	Telemetry Discrepancy	01/11/2024	01/11/2024
							F08	Hospitalization or Prolonged Hospitalization	A0712	Pacing Problem	01/11/2024	01/11/2024
									A0908	Incorrect, Inadequate or Imprecise Result or Readings	01/11/2024	01/11/2024
									A1304	Telemetry Discrepancy	01/11/2024	01/11/2024
							F1905	Device Revision or Replacement	A0712	Pacing Problem	01/11/2024	01/11/2024
									A0908	Incorrect, Inadequate or Imprecise Result or Readings	01/11/2024	01/11/2024
									A1304	Telemetry Discrepancy	01/11/2024	01/11/2024
16196	2024	All other reportable incidents	37265	Dual-chamber implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	18/11/2024	18/11/2024
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	18/11/2024	18/11/2024

							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	18/11/2024	18/11/2024
16197	2024		47265	Dual-chamber implantable pacemaker, rate-responsive	CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	29/07/2024	29/07/2024
							F1901	Additional Surgery	A24	No Apparent Device Problem	29/07/2024	29/07/2024
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	29/07/2024	29/07/2024
							F1901	Additional Surgery	A24	No Apparent Device Problem	29/07/2024	29/07/2024
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	29/07/2024	29/07/2024
							F1901	Additional Surgery	A24	No Apparent Device Problem	29/07/2024	29/07/2024
16198	2024	All other reportable incidents	47265	Dual-chamber implantable pacemaker, rate-responsive	CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	20/11/2024	20/11/2024
									A071102	Failure to Interrogate	20/11/2024	20/11/2024
							F1901	Additional Surgery	A070504	Premature Discharge of Battery	20/11/2024	20/11/2024
									A071102	Failure to Interrogate	20/11/2024	20/11/2024
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	20/11/2024	20/11/2024
									A071102	Failure to Interrogate	20/11/2024	20/11/2024
							F1901	Additional Surgery	A070504	Premature Discharge of Battery	20/11/2024	20/11/2024
									A071102	Failure to Interrogate	20/11/2024	20/11/2024
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	20/11/2024	20/11/2024
									A071102	Failure to Interrogate	20/11/2024	20/11/2024
							F1901	Additional Surgery	A070504	Premature Discharge of Battery	20/11/2024	20/11/2024
									A071102	Failure to Interrogate	20/11/2024	20/11/2024
16199	2024	All other reportable incidents	17882	Physiologic-monitoring defibrillation system	CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070802	Intermittent Loss of Power		
							F27	Problem Identified During Non-Clinical Procedure	A070802	Intermittent Loss of Power		
16200	2024		47265	Dual-chamber implantable pacemaker, rate-responsive	CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A070501	Battery Problem: High Impedance		
									A071205	Pacemaker Found in Back-Up Mode		
							F1905	Device Revision or Replacement	A070501	Battery Problem: High Impedance		
									A071205	Pacemaker Found in Back-Up Mode		
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A070501	Battery Problem: High Impedance		
									A071205	Pacemaker Found in Back-Up Mode		
							F1905	Device Revision or Replacement	A070501	Battery Problem: High Impedance		
									A071205	Pacemaker Found in Back-Up Mode		
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A070501	Battery Problem: High Impedance		
									A071205	Pacemaker Found in Back-Up Mode		
							F1905	Device Revision or Replacement	A070501	Battery Problem: High Impedance		
									A071205	Pacemaker Found in Back-Up Mode		
16201	2024	All other reportable incidents	47270	Cardiac resynchronization therapy implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	30/01/2025	30/01/2025
									A0908	Incorrect, Inadequate or Imprecise Result or Readings	30/01/2025	30/01/2025
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	30/01/2025	30/01/2025
									A0908	Incorrect, Inadequate or Imprecise Result or Readings	30/01/2025	30/01/2025
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	30/01/2025	30/01/2025
									A0908	Incorrect, Inadequate or Imprecise Result or Readings	30/01/2025	30/01/2025
16202	2024		47265	Dual-chamber implantable pacemaker, rate-responsive	CT1685	Cardiac pacemakers	E2321	Low Blood Pressure/ Hypotension	A0705	Battery Problem		
									A071205	Pacemaker Found in Back-Up Mode		
							F1905	Device Revision or Replacement	A0705	Battery Problem		
									A071205	Pacemaker Found in Back-Up Mode		
					CT257	Implantable cardiac pacemakers	E2321	Low Blood Pressure/ Hypotension	A0705	Battery Problem		
									A071205	Pacemaker Found in Back-Up Mode		
							F1905	Device Revision or Replacement	A0705	Battery Problem		
									A071205	Pacemaker Found in Back-Up Mode		
					CT2866	Implantable cardiac electrical stimulators	E2321	Low Blood Pressure/ Hypotension	A0705	Battery Problem		
									A071205	Pacemaker Found in Back-Up Mode		
							F1905	Device Revision or Replacement	A0705	Battery Problem		
									A071205	Pacemaker Found in Back-Up Mode		
16203	2024		47265	Dual-chamber implantable pacemaker, rate-responsive	CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	16/10/2024	16/10/2024
							F1901	Additional Surgery	A24	No Apparent Device Problem	16/10/2024	16/10/2024
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	16/10/2024	16/10/2024
							F1901	Additional Surgery	A24	No Apparent Device Problem	16/10/2024	16/10/2024
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	16/10/2024	16/10/2024
							F1901	Additional Surgery	A24	No Apparent Device Problem	16/10/2024	16/10/2024
16204	2024	All other reportable incidents	47270	Cardiac resynchronization therapy implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	07/11/2024	07/11/2024
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	07/11/2024	07/11/2024
							F12	Serious Injury/ Illness/ Impairment	A070504	Premature Discharge of Battery	07/11/2024	07/11/2024
							F1903	Device Explantation	A070504	Premature Discharge of Battery	07/11/2024	07/11/2024
16205	2024		47267	Single-chamber implantable pacemaker, rate-responsive	CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem		
									A071205	Pacemaker Found in Back-Up Mode		
							F1905	Device Revision or Replacement	A0705	Battery Problem		
									A071205	Pacemaker Found in Back-Up Mode		
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem		
									A071205	Pacemaker Found in Back-Up Mode		
							F1905	Device Revision or Replacement	A0705	Battery Problem		
									A071205	Pacemaker Found in Back-Up Mode		
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem		
									A071205	Pacemaker Found in Back-Up Mode		
							F1905	Device Revision or Replacement	A0705	Battery Problem		
									A071205	Pacemaker Found in Back-Up Mode		
16206	2024	All other reportable incidents	37265	Dual-chamber implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A072001	Electromagnetic Interference		
							F26	No Health Consequences or Impact	A072001	Electromagnetic Interference		
16207	2024	All other reportable incidents	47263	Cardiac resynchronization therapy implantable pacemaker	CT1685	Cardiac pacemakers	E011903	Syncope/Fainting	A070909	Oversensing	12/11/2024	12/11/2024
									A0908	Incorrect, Inadequate or Imprecise Result or Readings	12/11/2024	12/11/2024
									A1304	Telemetry Discrepancy	12/11/2024	12/11/2024
							F08	Hospitalization or Prolonged Hospitalization	A070909	Oversensing	12/11/2024	12/11/2024
									A0908	Incorrect, Inadequate or Imprecise Result or Readings	12/11/2024	12/11/2024
									A1304	Telemetry Discrepancy	12/11/2024	12/11/2024
							F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A070909	Oversensing	12/11/2024	12/11/2024
									A0908	Incorrect, Inadequate or Imprecise Result or Readings	12/11/2024	12/11/2024

								A1304	Telemetry Discrepancy	12/11/2024	12/11/2024
						F1905	Device Revision or Replacement	A070909	Oversensing	12/11/2024	12/11/2024
								A0908	Incorrect, Inadequate or Imprecise Result or Readings	12/11/2024	12/11/2024
								A1304	Telemetry Discrepancy	12/11/2024	12/11/2024
				CT257	Implantable cardiac pacemakers	E011903	Syncope/Fainting	A070909	Oversensing	12/11/2024	12/11/2024
								A0908	Incorrect, Inadequate or Imprecise Result or Readings	12/11/2024	12/11/2024
								A1304	Telemetry Discrepancy	12/11/2024	12/11/2024
						F08	Hospitalization or Prolonged Hospitalization	A070909	Oversensing	12/11/2024	12/11/2024
								A0908	Incorrect, Inadequate or Imprecise Result or Readings	12/11/2024	12/11/2024
								A1304	Telemetry Discrepancy	12/11/2024	12/11/2024
						F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A070909	Oversensing	12/11/2024	12/11/2024
								A0908	Incorrect, Inadequate or Imprecise Result or Readings	12/11/2024	12/11/2024
								A1304	Telemetry Discrepancy	12/11/2024	12/11/2024
						F1905	Device Revision or Replacement	A070909	Oversensing	12/11/2024	12/11/2024
								A0908	Incorrect, Inadequate or Imprecise Result or Readings	12/11/2024	12/11/2024
								A1304	Telemetry Discrepancy	12/11/2024	12/11/2024
				CT2866	Implantable cardiac electrical stimulators	E011903	Syncope/Fainting	A070909	Oversensing	12/11/2024	12/11/2024
								A0908	Incorrect, Inadequate or Imprecise Result or Readings	12/11/2024	12/11/2024
								A1304	Telemetry Discrepancy	12/11/2024	12/11/2024
						F08	Hospitalization or Prolonged Hospitalization	A070909	Oversensing	12/11/2024	12/11/2024
								A0908	Incorrect, Inadequate or Imprecise Result or Readings	12/11/2024	12/11/2024
								A1304	Telemetry Discrepancy	12/11/2024	12/11/2024
						F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A070909	Oversensing	12/11/2024	12/11/2024
								A0908	Incorrect, Inadequate or Imprecise Result or Readings	12/11/2024	12/11/2024
								A1304	Telemetry Discrepancy	12/11/2024	12/11/2024
						F1905	Device Revision or Replacement	A070909	Oversensing	12/11/2024	12/11/2024
								A0908	Incorrect, Inadequate or Imprecise Result or Readings	12/11/2024	12/11/2024
								A1304	Telemetry Discrepancy	12/11/2024	12/11/2024
16208	2024	All other reportable incidents	47267	Single-chamber implantable pacemaker, rate-responsive	CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
16209	2024	All other reportable incidents	47265	Dual-chamber implantable pacemaker, rate-responsive	CT1685	Cardiac pacemakers	E0112	Dizziness	A070504	Premature Discharge of Battery	10/10/2024	10/10/2024
									A071102	Failure to Interrogate	10/10/2024	10/10/2024
									A071204	No Pacing	10/10/2024	10/10/2024
							E0601	Arrhythmia	A070504	Premature Discharge of Battery	10/10/2024	10/10/2024
									A071102	Failure to Interrogate	10/10/2024	10/10/2024
									A071204	No Pacing	10/10/2024	10/10/2024
							E2312	Fatigue	A070504	Premature Discharge of Battery	10/10/2024	10/10/2024
									A071102	Failure to Interrogate	10/10/2024	10/10/2024
									A071204	No Pacing	10/10/2024	10/10/2024
							F1901	Additional Surgery	A070504	Premature Discharge of Battery	10/10/2024	10/10/2024
									A071102	Failure to Interrogate	10/10/2024	10/10/2024
									A071204	No Pacing	10/10/2024	10/10/2024
					CT257	Implantable cardiac pacemakers	E0112	Dizziness	A070504	Premature Discharge of Battery	10/10/2024	10/10/2024
									A071102	Failure to Interrogate	10/10/2024	10/10/2024
									A071204	No Pacing	10/10/2024	10/10/2024
							E0601	Arrhythmia	A070504	Premature Discharge of Battery	10/10/2024	10/10/2024
									A071102	Failure to Interrogate	10/10/2024	10/10/2024
									A071204	No Pacing	10/10/2024	10/10/2024
							E2312	Fatigue	A070504	Premature Discharge of Battery	10/10/2024	10/10/2024
									A071102	Failure to Interrogate	10/10/2024	10/10/2024
									A071204	No Pacing	10/10/2024	10/10/2024
							F1901	Additional Surgery	A070504	Premature Discharge of Battery	10/10/2024	10/10/2024
									A071102	Failure to Interrogate	10/10/2024	10/10/2024
									A071204	No Pacing	10/10/2024	10/10/2024
					CT2866	Implantable cardiac electrical stimulators	E0112	Dizziness	A070504	Premature Discharge of Battery	10/10/2024	10/10/2024
									A071102	Failure to Interrogate	10/10/2024	10/10/2024
									A071204	No Pacing	10/10/2024	10/10/2024
							E0601	Arrhythmia	A070504	Premature Discharge of Battery	10/10/2024	10/10/2024
									A071102	Failure to Interrogate	10/10/2024	10/10/2024
									A071204	No Pacing	10/10/2024	10/10/2024
							E2312	Fatigue	A070504	Premature Discharge of Battery	10/10/2024	10/10/2024
									A071102	Failure to Interrogate	10/10/2024	10/10/2024
									A071204	No Pacing	10/10/2024	10/10/2024
							F1901	Additional Surgery	A070504	Premature Discharge of Battery	10/10/2024	10/10/2024
									A071102	Failure to Interrogate	10/10/2024	10/10/2024
									A071204	No Pacing	10/10/2024	10/10/2024
16210	2024	Death	37265	Dual-chamber implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem		
									A24	No Apparent Device Problem		
							F26	No Health Consequences or Impact	A0705	Battery Problem		
									A24	No Apparent Device Problem		
16211	2024	All other reportable incidents	35852	Single-chamber implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem	04/11/2024	04/11/2024
									A070504	Premature Discharge of Battery	04/11/2024	04/11/2024
							F08	Hospitalization or Prolonged Hospitalization	A0705	Battery Problem	04/11/2024	04/11/2024
									A070504	Premature Discharge of Battery	04/11/2024	04/11/2024
							F1905	Device Revision or Replacement	A0705	Battery Problem	04/11/2024	04/11/2024
									A070504	Premature Discharge of Battery	04/11/2024	04/11/2024
16212	2024	All other reportable incidents	17882	Physiologic-monitoring defibrillation system	CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070912	Incorrect Interpretation of Signal		

							F26	No Health Consequences or Impact	A070912	Incorrect Interpretation of Signal		
16213	2024	All other reportable incidents	17882	Physiologic-monitoring defibrillation system	CT1684	Cardiac pacemakers and associated devices	E0623	Unspecified Heart Problem	A071301	Failure to Deliver Shock/Stimulation		
							F05	Delay to Treatment/ Therapy	A071301	Failure to Deliver Shock/Stimulation		
16214	2024	All other reportable incidents	37265	Dual-chamber implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A160602	Premature End-of-Life Indicator		
							F1905	Device Revision or Replacement	A160602	Premature End-of-Life Indicator		
16215	2024	All other reportable incidents	17882	Physiologic-monitoring defibrillation system	CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070908	Failure to Sense		
							F27	Problem Identified During Non-Clinical Procedure	A070908	Failure to Sense		
16216	2024	All other reportable incidents	37265	Dual-chamber implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	25/11/2024	25/11/2024
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	25/11/2024	25/11/2024
							F12	Serious Injury/ Illness/ Impairment	A070504	Premature Discharge of Battery	25/11/2024	25/11/2024
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	25/11/2024	25/11/2024
16217	2024		35852	Single-chamber implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A090402	Failure to Deliver Energy		
									A26	Insufficient Device Problem Information		
							F28	Appropriate Health Impact Term/Code not Available	A090402	Failure to Deliver Energy		
									A26	Insufficient Device Problem Information		
16218	2024	All other reportable incidents	35852	Single-chamber implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E060109	Tachycardia	A071302	Inappropriate/Inadequate Shock/Stimulation		
							E210401	Shock from Patient Lead(s)	A071302	Inappropriate/Inadequate Shock/Stimulation		
							F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A071302	Inappropriate/Inadequate Shock/Stimulation		
							F1203	Life Threatening Illness or Injury	A071302	Inappropriate/Inadequate Shock/Stimulation		
16219	2024		47265	Dual-chamber implantable pacemaker, rate-responsive	CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	02/11/2024	02/11/2024
							F1901	Additional Surgery	A24	No Apparent Device Problem	02/11/2024	02/11/2024
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	02/11/2024	02/11/2024
							F1901	Additional Surgery	A24	No Apparent Device Problem	02/11/2024	02/11/2024
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	02/11/2024	02/11/2024
							F1901	Additional Surgery	A24	No Apparent Device Problem	02/11/2024	02/11/2024
16220	2024	All other reportable incidents	35852	Single-chamber implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	26/03/2025	26/03/2025
									A0908	Incorrect, Inadequate or Imprecise Result or Readings	26/03/2025	26/03/2025
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	26/03/2025	26/03/2025
									A0908	Incorrect, Inadequate or Imprecise Result or Readings	26/03/2025	26/03/2025
							F1903	Device Explantation	A070504	Premature Discharge of Battery	26/03/2025	26/03/2025
									A0908	Incorrect, Inadequate or Imprecise Result or Readings	26/03/2025	26/03/2025
16221	2024		44862	Temporary cardiac pacing/stimulation connection cable, reusable	CT1684	Cardiac pacemakers and associated devices	E2401	Insufficient Information	A0508	Noise, Audible		
							F1001	Absence of Treatment	A0508	Noise, Audible		
16222	2024	All other reportable incidents	17882	Physiologic-monitoring defibrillation system	CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A0401	Break		
							F27	Problem Identified During Non-Clinical Procedure	A0401	Break		
16223	2024	All other reportable incidents	47265	Dual-chamber implantable pacemaker, rate-responsive	CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A13	Communication or Transmission Problem	06/09/2024	06/09/2024
							F26	No Health Consequences or Impact	A13	Communication or Transmission Problem	06/09/2024	06/09/2024
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A13	Communication or Transmission Problem	06/09/2024	06/09/2024
							F26	No Health Consequences or Impact	A13	Communication or Transmission Problem	06/09/2024	06/09/2024
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A13	Communication or Transmission Problem	06/09/2024	06/09/2024
							F26	No Health Consequences or Impact	A13	Communication or Transmission Problem	06/09/2024	06/09/2024
16224	2024	All other reportable incidents	47265	Dual-chamber implantable pacemaker, rate-responsive	CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A0908	Incorrect, Inadequate or Imprecise Result or Readings	21/05/2024	21/05/2024
							F08	Hospitalization or Prolonged Hospitalization	A0908	Incorrect, Inadequate or Imprecise Result or Readings	21/05/2024	21/05/2024
							F1905	Device Revision or Replacement	A0908	Incorrect, Inadequate or Imprecise Result or Readings	21/05/2024	21/05/2024
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A0908	Incorrect, Inadequate or Imprecise Result or Readings	21/05/2024	21/05/2024
							F08	Hospitalization or Prolonged Hospitalization	A0908	Incorrect, Inadequate or Imprecise Result or Readings	21/05/2024	21/05/2024
							F1905	Device Revision or Replacement	A0908	Incorrect, Inadequate or Imprecise Result or Readings	21/05/2024	21/05/2024
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A0908	Incorrect, Inadequate or Imprecise Result or Readings	21/05/2024	21/05/2024
							F08	Hospitalization or Prolonged Hospitalization	A0908	Incorrect, Inadequate or Imprecise Result or Readings	21/05/2024	21/05/2024
							F1905	Device Revision or Replacement	A0908	Incorrect, Inadequate or Imprecise Result or Readings	21/05/2024	21/05/2024
16225	2024		47265	Dual-chamber implantable pacemaker, rate-responsive	CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem		
							F1901	Additional Surgery	A24	No Apparent Device Problem		
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem		
							F1901	Additional Surgery	A24	No Apparent Device Problem		
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem		
							F1901	Additional Surgery	A24	No Apparent Device Problem		
16226	2024		47265	Dual-chamber implantable pacemaker, rate-responsive	CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	16/10/2024	16/10/2024
							F1901	Additional Surgery	A24	No Apparent Device Problem	16/10/2024	16/10/2024
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	16/10/2024	16/10/2024
							F1901	Additional Surgery	A24	No Apparent Device Problem	16/10/2024	16/10/2024
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	16/10/2024	16/10/2024
							F1901	Additional Surgery	A24	No Apparent Device Problem	16/10/2024	16/10/2024
16227	2024		47265	Dual-chamber implantable pacemaker, rate-responsive	CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A0701	Capturing Problem		
									A070910	Undersensing		
							F26	No Health Consequences or Impact	A0701	Capturing Problem		
									A070910	Undersensing		
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A0701	Capturing Problem		
									A070910	Undersensing		
							F26	No Health Consequences or Impact	A0701	Capturing Problem		
									A070910	Undersensing		
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A0701	Capturing Problem		
									A070910	Undersensing		
							F26	No Health Consequences or Impact	A0701	Capturing Problem		
									A070910	Undersensing		
16228	2024	All other reportable incidents	47270	Cardiac resynchronization therapy implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	19/11/2024	19/11/2024
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	19/11/2024	19/11/2024
							F12	Serious Injury/ Illness/ Impairment	A070504	Premature Discharge of Battery	19/11/2024	19/11/2024
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	19/11/2024	19/11/2024
16229	2024	All other reportable incidents	47267	Single-chamber implantable pacemaker, rate-	CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A070909	Oversensing		

				responsive					A071204	No Pacing		
							F22	Unexpected Diagnostic Intervention	A070909	Oversensing		
									A071204	No Pacing		
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A070909	Oversensing		
									A071204	No Pacing		
							F22	Unexpected Diagnostic Intervention	A070909	Oversensing		
									A071204	No Pacing		
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A070909	Oversensing		
									A071204	No Pacing		
							F22	Unexpected Diagnostic Intervention	A070909	Oversensing		
									A071204	No Pacing		
16230	2024	All other reportable incidents	47265	Dual-chamber implantable pacemaker, rate-responsive	CT1685	Cardiac pacemakers	E060109	Tachycardia	A0104	Migration or Expulsion of Device		
							E233001	Chest Pain	A0104	Migration or Expulsion of Device		
							F12	Serious Injury/ Illness/ Impairment	A0104	Migration or Expulsion of Device		
							F2303	Medication Required	A0104	Migration or Expulsion of Device		
					CT257	Implantable cardiac pacemakers	E060109	Tachycardia	A0104	Migration or Expulsion of Device		
							E233001	Chest Pain	A0104	Migration or Expulsion of Device		
							F12	Serious Injury/ Illness/ Impairment	A0104	Migration or Expulsion of Device		
							F2303	Medication Required	A0104	Migration or Expulsion of Device		
					CT2866	Implantable cardiac electrical stimulators	E060109	Tachycardia	A0104	Migration or Expulsion of Device		
							E233001	Chest Pain	A0104	Migration or Expulsion of Device		
							F12	Serious Injury/ Illness/ Impairment	A0104	Migration or Expulsion of Device		
							F2303	Medication Required	A0104	Migration or Expulsion of Device		
16231	2024	All other reportable incidents	35852	Single-chamber implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	03/02/2025	03/02/2025
									A0908	Incorrect, Inadequate or Imprecise Result or Readings	03/02/2025	03/02/2025
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	03/02/2025	03/02/2025
									A0908	Incorrect, Inadequate or Imprecise Result or Readings	03/02/2025	03/02/2025
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	03/02/2025	03/02/2025
									A0908	Incorrect, Inadequate or Imprecise Result or Readings	03/02/2025	03/02/2025
16232	2024	All other reportable incidents	37265	Dual-chamber implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A072201	High Impedance		
									A1107	Data Problem		
							F26	No Health Consequences or Impact	A072201	High Impedance		
									A1107	Data Problem		
16233	2024	All other reportable incidents	47270	Cardiac resynchronization therapy implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	01/01/2024	31/12/2024
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	01/01/2024	31/12/2024
							F12	Serious Injury/ Illness/ Impairment	A070504	Premature Discharge of Battery	01/01/2024	31/12/2024
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	01/01/2024	31/12/2024
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery	01/01/2024	31/12/2024
16234	2024	All other reportable incidents	60789	Intracardiac pacemaker	CT1685	Cardiac pacemakers	E0604	Cardiac Perforation	A0701	Capturing Problem		
							E0619	Pericardial Effusion	A0701	Capturing Problem		
							F12	Serious Injury/ Illness/ Impairment	A0701	Capturing Problem		
					CT257	Implantable cardiac pacemakers	E0604	Cardiac Perforation	A0701	Capturing Problem		
							E0619	Pericardial Effusion	A0701	Capturing Problem		
							F12	Serious Injury/ Illness/ Impairment	A0701	Capturing Problem		
					CT2818	Catheterization/Cannulation/Fluid path devices	E0604	Cardiac Perforation	A0701	Capturing Problem		
							E0619	Pericardial Effusion	A0701	Capturing Problem		
							F12	Serious Injury/ Illness/ Impairment	A0701	Capturing Problem		
					CT2866	Implantable cardiac electrical stimulators	E0604	Cardiac Perforation	A0701	Capturing Problem		
							E0619	Pericardial Effusion	A0701	Capturing Problem		
							F12	Serious Injury/ Illness/ Impairment	A0701	Capturing Problem		
16235	2024	All other reportable incidents	47205	Cardiac pulse generator programmer	CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A0902	Display or Visual Feedback Problem		
							F27	Problem Identified During Non-Clinical Procedure	A0902	Display or Visual Feedback Problem		
16236	2024	All other reportable incidents	47265	Dual-chamber implantable pacemaker, rate-responsive	CT1685	Cardiac pacemakers	E0717	Dyspnea	A0712	Pacing Problem		
							E2401	Insufficient Information	A0712	Pacing Problem		
							F24	Insufficient Health Impact Information	A0712	Pacing Problem		
					CT257	Implantable cardiac pacemakers	E0717	Dyspnea	A0712	Pacing Problem		
							E2401	Insufficient Information	A0712	Pacing Problem		
							F24	Insufficient Health Impact Information	A0712	Pacing Problem		
					CT2866	Implantable cardiac electrical stimulators	E0717	Dyspnea	A0712	Pacing Problem		
							E2401	Insufficient Information	A0712	Pacing Problem		
							F24	Insufficient Health Impact Information	A0712	Pacing Problem		
16237	2024	All other reportable incidents	47270	Cardiac resynchronization therapy implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A0711	Interrogation Problem	19/11/2024	19/11/2024
							F08	Hospitalization or Prolonged Hospitalization	A0711	Interrogation Problem	19/11/2024	19/11/2024
							F1905	Device Revision or Replacement	A0711	Interrogation Problem	19/11/2024	19/11/2024
16238	2024	All other reportable incidents	37265	Dual-chamber implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	03/02/2025	03/02/2025
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	03/02/2025	03/02/2025
							F1903	Device Explantation	A070504	Premature Discharge of Battery	03/02/2025	03/02/2025
16239	2024		47265	Dual-chamber implantable pacemaker, rate-responsive	CT1685	Cardiac pacemakers	E2401	Insufficient Information	A0711	Interrogation Problem		
									A071102	Failure to Interrogate		
							F02	Death	A0711	Interrogation Problem		
									A071102	Failure to Interrogate		
					CT257	Implantable cardiac pacemakers	E2401	Insufficient Information	A0711	Interrogation Problem		
									A071102	Failure to Interrogate		
							F02	Death	A0711	Interrogation Problem		
									A071102	Failure to Interrogate		
					CT2866	Implantable cardiac electrical stimulators	E2401	Insufficient Information	A0711	Interrogation Problem		
									A071102	Failure to Interrogate		
							F02	Death	A0711	Interrogation Problem		
									A071102	Failure to Interrogate		
16240	2024		47265	Dual-chamber implantable pacemaker, rate-responsive	CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem		
							F1901	Additional Surgery	A24	No Apparent Device Problem		
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem		
							F1901	Additional Surgery	A24	No Apparent Device Problem		
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem		
							F1901	Additional Surgery	A24	No Apparent Device Problem		

16241	2024	All other reportable incidents	47265	Dual-chamber implantable pacemaker, rate-responsive	CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
									A071102	Failure to Interrogate		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
									A071102	Failure to Interrogate		
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
									A071102	Failure to Interrogate		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
									A071102	Failure to Interrogate		
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
									A071102	Failure to Interrogate		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
									A071102	Failure to Interrogate		
16242	2024	All other reportable incidents	17882	Physiologic-monitoring defibrillation system	CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A0719	Unexpected Shutdown		
							F27	Problem Identified During Non-Clinical Procedure	A0719	Unexpected Shutdown		
16243	2024	Unanticipated serious deterioration in state of health	60789	Intracardiac pacemaker	CT1685	Cardiac pacemakers	E0604	Cardiac Perforation	A24	No Apparent Device Problem		
							E0619	Pericardial Effusion	A24	No Apparent Device Problem		
							E1020	Nausea	A24	No Apparent Device Problem		
							E233001	Chest Pain	A24	No Apparent Device Problem		
							F08	Hospitalization or Prolonged Hospitalization	A24	No Apparent Device Problem		
							F1203	Life Threatening Illness or Injury	A24	No Apparent Device Problem		
							F1901	Additional Surgery	A24	No Apparent Device Problem		
					CT257	Implantable cardiac pacemakers	E0604	Cardiac Perforation	A24	No Apparent Device Problem		
							E0619	Pericardial Effusion	A24	No Apparent Device Problem		
							E1020	Nausea	A24	No Apparent Device Problem		
							E233001	Chest Pain	A24	No Apparent Device Problem		
							F08	Hospitalization or Prolonged Hospitalization	A24	No Apparent Device Problem		
							F1203	Life Threatening Illness or Injury	A24	No Apparent Device Problem		
							F1901	Additional Surgery	A24	No Apparent Device Problem		
					CT2818	Catheterization/Cannulation/Fluid path devices	E0604	Cardiac Perforation	A24	No Apparent Device Problem		
							E0619	Pericardial Effusion	A24	No Apparent Device Problem		
							E1020	Nausea	A24	No Apparent Device Problem		
							E233001	Chest Pain	A24	No Apparent Device Problem		
							F08	Hospitalization or Prolonged Hospitalization	A24	No Apparent Device Problem		
							F1203	Life Threatening Illness or Injury	A24	No Apparent Device Problem		
							F1901	Additional Surgery	A24	No Apparent Device Problem		
					CT2866	Implantable cardiac electrical stimulators	E0604	Cardiac Perforation	A24	No Apparent Device Problem		
							E0619	Pericardial Effusion	A24	No Apparent Device Problem		
							E1020	Nausea	A24	No Apparent Device Problem		
							E233001	Chest Pain	A24	No Apparent Device Problem		
							F08	Hospitalization or Prolonged Hospitalization	A24	No Apparent Device Problem		
							F1203	Life Threatening Illness or Injury	A24	No Apparent Device Problem		
							F1901	Additional Surgery	A24	No Apparent Device Problem		
16244	2024	All other reportable incidents	47265	Dual-chamber implantable pacemaker, rate-responsive	CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A0203	Defective Device	01/11/2024	01/11/2024
									A24	No Apparent Device Problem	01/11/2024	01/11/2024
							F08	Hospitalization or Prolonged Hospitalization	A0203	Defective Device	01/11/2024	01/11/2024
									A24	No Apparent Device Problem	01/11/2024	01/11/2024
							F1905	Device Revision or Replacement	A0203	Defective Device	01/11/2024	01/11/2024
									A24	No Apparent Device Problem	01/11/2024	01/11/2024
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A0203	Defective Device	01/11/2024	01/11/2024
									A24	No Apparent Device Problem	01/11/2024	01/11/2024
							F08	Hospitalization or Prolonged Hospitalization	A0203	Defective Device	01/11/2024	01/11/2024
									A24	No Apparent Device Problem	01/11/2024	01/11/2024
							F1905	Device Revision or Replacement	A0203	Defective Device	01/11/2024	01/11/2024
									A24	No Apparent Device Problem	01/11/2024	01/11/2024
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A0203	Defective Device	01/11/2024	01/11/2024
									A24	No Apparent Device Problem	01/11/2024	01/11/2024
							F08	Hospitalization or Prolonged Hospitalization	A0203	Defective Device	01/11/2024	01/11/2024
									A24	No Apparent Device Problem	01/11/2024	01/11/2024
							F1905	Device Revision or Replacement	A0203	Defective Device	01/11/2024	01/11/2024
									A24	No Apparent Device Problem	01/11/2024	01/11/2024
16245	2024	All other reportable incidents	35852	Single-chamber implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E210401	Shock from Patient Lead(s)	A071302	Inappropriate/Inadequate Shock/Stimulation		
							F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A071302	Inappropriate/Inadequate Shock/Stimulation		
16246	2024	All other reportable incidents	47265	Dual-chamber implantable pacemaker, rate-responsive	CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A070909	Oversensing		
									A070910	Undersensing		
									A090801	Signal Artifact/Noise		
							F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A070909	Oversensing		
									A070910	Undersensing		
									A090801	Signal Artifact/Noise		
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A070909	Oversensing		
									A070910	Undersensing		
									A090801	Signal Artifact/Noise		
							F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A070909	Oversensing		
									A070910	Undersensing		
									A090801	Signal Artifact/Noise		
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A070909	Oversensing		
									A070910	Undersensing		
									A090801	Signal Artifact/Noise		
							F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A070909	Oversensing		
									A070910	Undersensing		
									A090801	Signal Artifact/Noise		
16247	2024	All other reportable incidents	47263	Cardiac resynchronization therapy implantable pacemaker	CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	16/01/2025	16/01/2025
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	16/01/2025	16/01/2025
							F12	Serious Injury/ Illness/ Impairment	A070504	Premature Discharge of Battery	16/01/2025	16/01/2025
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	16/01/2025	16/01/2025
							F2307	Reprogramming of Device	A070504	Premature Discharge of Battery	16/01/2025	16/01/2025
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	16/01/2025	16/01/2025
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	16/01/2025	16/01/2025
							F12	Serious Injury/ Illness/ Impairment	A070504	Premature Discharge of Battery	16/01/2025	16/01/2025

							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	16/01/2025	16/01/2025
							F2307	Reprogramming of Device	A070504	Premature Discharge of Battery	16/01/2025	16/01/2025
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	16/01/2025	16/01/2025
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	16/01/2025	16/01/2025
							F12	Serious Injury/ Illness/ Impairment	A070504	Premature Discharge of Battery	16/01/2025	16/01/2025
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	16/01/2025	16/01/2025
							F2307	Reprogramming of Device	A070504	Premature Discharge of Battery	16/01/2025	16/01/2025
16248	2024	All other reportable incidents	47265	Dual-chamber implantable pacemaker, rate-responsive	CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
16249	2024	All other reportable incidents	37265	Dual-chamber implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A1304	Telemetry Discrepancy	19/10/2023	19/10/2023
							F08	Hospitalization or Prolonged Hospitalization	A1304	Telemetry Discrepancy	19/10/2023	19/10/2023
							F1905	Device Revision or Replacement	A1304	Telemetry Discrepancy	19/10/2023	19/10/2023
16250	2024	All other reportable incidents	47270	Cardiac resynchronization therapy implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	28/02/2025	28/02/2025
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	28/02/2025	28/02/2025
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	28/02/2025	28/02/2025
16251	2024	All other reportable incidents	17882	Physiologic-monitoring defibrillation system	CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A0719	Unexpected Shutdown		
							F26	No Health Consequences or Impact	A0719	Unexpected Shutdown		
16252	2024	All other reportable incidents	35852	Single-chamber implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A0709	Device Sensing Problem		
									A0712	Pacing Problem		
							F26	No Health Consequences or Impact	A0709	Device Sensing Problem		
									A0712	Pacing Problem		
16253	2024	All other reportable incidents	17882	Physiologic-monitoring defibrillation system	CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A0801	Failure to Calibrate		
							F27	Problem Identified During Non-Clinical Procedure	A0801	Failure to Calibrate		
16254	2024	All other reportable incidents	47205	Cardiac pulse generator programmer	CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A13	Communication or Transmission Problem		
							F23	Unexpected Medical Intervention	A13	Communication or Transmission Problem		
							F26	No Health Consequences or Impact	A13	Communication or Transmission Problem		
16255	2024	All other reportable incidents	17882	Physiologic-monitoring defibrillation system	CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070802	Intermittent Loss of Power		
							F27	Problem Identified During Non-Clinical Procedure	A070802	Intermittent Loss of Power		
16256	2024	All other reportable incidents	47265	Dual-chamber implantable pacemaker, rate-responsive	CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
16257	2024	All other reportable incidents	47263	Cardiac resynchronization therapy implantable pacemaker	CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A070909	Oversensing	21/07/2025	21/07/2025
							F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A070909	Oversensing	21/07/2025	21/07/2025
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A070909	Oversensing	21/07/2025	21/07/2025
							F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A070909	Oversensing	21/07/2025	21/07/2025
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A070909	Oversensing	21/07/2025	21/07/2025
							F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A070909	Oversensing	21/07/2025	21/07/2025
16258	2024	All other reportable incidents			CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A0203	Defective Device	16/02/2024	16/02/2024
							F26	No Health Consequences or Impact	A0203	Defective Device	16/02/2024	16/02/2024
16259	2024	All other reportable incidents			CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070101	Failure to Capture		
							F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A070101	Failure to Capture		
16260	2024	All other reportable incidents	17882	Physiologic-monitoring defibrillation system	CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070803	Failure to Power Up		
							F27	Problem Identified During Non-Clinical Procedure	A070803	Failure to Power Up		
16261	2024	All other reportable incidents	36106	Implantable pacemaker analyser, intraoperative	CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A071102	Failure to Interrogate		
									A13	Communication or Transmission Problem		
							F26	No Health Consequences or Impact	A071102	Failure to Interrogate		
									A13	Communication or Transmission Problem		
16262	2024	All other reportable incidents	47265	Dual-chamber implantable pacemaker, rate-responsive	CT1685	Cardiac pacemakers	E011903	Syncope/Fainting	A0908	Incorrect, Inadequate or Imprecise Result or Readings	06/09/2024	06/09/2024
							E060101	Asystole	A0908	Incorrect, Inadequate or Imprecise Result or Readings	06/09/2024	06/09/2024
							F08	Hospitalization or Prolonged Hospitalization	A0908	Incorrect, Inadequate or Imprecise Result or Readings	06/09/2024	06/09/2024
							F1905	Device Revision or Replacement	A0908	Incorrect, Inadequate or Imprecise Result or Readings	06/09/2024	06/09/2024
							F2303	Medication Required	A0908	Incorrect, Inadequate or Imprecise Result or Readings	06/09/2024	06/09/2024
					CT257	Implantable cardiac pacemakers	E011903	Syncope/Fainting	A0908	Incorrect, Inadequate or Imprecise Result or Readings	06/09/2024	06/09/2024
							E060101	Asystole	A0908	Incorrect, Inadequate or Imprecise Result or Readings	06/09/2024	06/09/2024
							F08	Hospitalization or Prolonged Hospitalization	A0908	Incorrect, Inadequate or Imprecise Result or Readings	06/09/2024	06/09/2024
							F1905	Device Revision or Replacement	A0908	Incorrect, Inadequate or Imprecise Result or Readings	06/09/2024	06/09/2024
							F2303	Medication Required	A0908	Incorrect, Inadequate or Imprecise Result or Readings	06/09/2024	06/09/2024
					CT2866	Implantable cardiac electrical stimulators	E011903	Syncope/Fainting	A0908	Incorrect, Inadequate or Imprecise Result or Readings	06/09/2024	06/09/2024
							E060101	Asystole	A0908	Incorrect, Inadequate or Imprecise Result or Readings	06/09/2024	06/09/2024
							F08	Hospitalization or Prolonged Hospitalization	A0908	Incorrect, Inadequate or Imprecise Result or Readings	06/09/2024	06/09/2024
							F1905	Device Revision or Replacement	A0908	Incorrect, Inadequate or Imprecise Result or Readings	06/09/2024	06/09/2024
							F2303	Medication Required	A0908	Incorrect, Inadequate or Imprecise Result or Readings	06/09/2024	06/09/2024
16263	2024	All other reportable incidents	47265	Dual-chamber implantable pacemaker, rate-	CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A070909	Oversensing		

				responsive					A0712	Pacing Problem		
							F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A070909	Oversensing		
									A0712	Pacing Problem		
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A070909	Oversensing		
									A0712	Pacing Problem		
							F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A070909	Oversensing		
									A0712	Pacing Problem		
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A070909	Oversensing		
									A0712	Pacing Problem		
							F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A070909	Oversensing		
									A0712	Pacing Problem		
16264	2024	All other reportable incidents	35852	Single-chamber implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E210401	Shock from Patient Lead(s)	A0709	Device Sensing Problem		
									A0712	Pacing Problem		
									A071302	Inappropriate/Inadequate Shock/Stimulation		
							F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A0709	Device Sensing Problem		
									A0712	Pacing Problem		
									A071302	Inappropriate/Inadequate Shock/Stimulation		
							F1203	Life Threatening Illness or Injury	A0709	Device Sensing Problem		
									A0712	Pacing Problem		
									A071302	Inappropriate/Inadequate Shock/Stimulation		
16265	2024	All other reportable incidents	37265	Dual-chamber implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A05	Mechanical Problem		
									A072201	High Impedance		
							F26	No Health Consequences or Impact	A05	Mechanical Problem		
									A072201	High Impedance		
16266	2024	All other reportable incidents	47270	Cardiac resynchronization therapy implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E0601	Arrhythmia	A070912	Incorrect Interpretation of Signal		
									A071201	Failure to Convert Rhythm		
									A071302	Inappropriate/Inadequate Shock/Stimulation		
							F12	Serious Injury/ Illness/ Impairment	A070912	Incorrect Interpretation of Signal		
									A071201	Failure to Convert Rhythm		
									A071302	Inappropriate/Inadequate Shock/Stimulation		
16267	2024		47265	Dual-chamber implantable pacemaker, rate-responsive	CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A070909	Oversensing	23/08/2024	23/08/2024
									A1205	Loose or Intermittent Connection	23/08/2024	23/08/2024
							F1901	Additional Surgery	A070909	Oversensing	23/08/2024	23/08/2024
									A1205	Loose or Intermittent Connection	23/08/2024	23/08/2024
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A070909	Oversensing	23/08/2024	23/08/2024
									A1205	Loose or Intermittent Connection	23/08/2024	23/08/2024
							F1901	Additional Surgery	A070909	Oversensing	23/08/2024	23/08/2024
									A1205	Loose or Intermittent Connection	23/08/2024	23/08/2024
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A070909	Oversensing	23/08/2024	23/08/2024
									A1205	Loose or Intermittent Connection	23/08/2024	23/08/2024
							F1901	Additional Surgery	A070909	Oversensing	23/08/2024	23/08/2024
									A1205	Loose or Intermittent Connection	23/08/2024	23/08/2024
16268	2024	All other reportable incidents	35852	Single-chamber implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	26/11/2024	26/11/2024
							F1901	Additional Surgery	A070504	Premature Discharge of Battery	26/11/2024	26/11/2024
16269	2024	All other reportable incidents	35852	Single-chamber implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
16270	2024	All other reportable incidents	17882	Physiologic-monitoring defibrillation system	CT1684	Cardiac pacemakers and associated devices	E0602	Cardiac Arrest	A090201	Device Displays Incorrect Message		
							F05	Delay to Treatment/ Therapy	A090201	Device Displays Incorrect Message		
16271	2024	All other reportable incidents	47206	Cardiac pulse generator software	CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A0203	Defective Device		
									A150105	Key or Button Unresponsive/not Working		
							F26	No Health Consequences or Impact	A0203	Defective Device		
									A150105	Key or Button Unresponsive/not Working		
16272	2024	All other reportable incidents	60789	Intracardiac pacemaker	CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem		
							F26	No Health Consequences or Impact	A24	No Apparent Device Problem		
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem		
							F26	No Health Consequences or Impact	A24	No Apparent Device Problem		
					CT2818	Catheterization/Cannulation/Fluid path devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem		
							F26	No Health Consequences or Impact	A24	No Apparent Device Problem		
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem		
							F26	No Health Consequences or Impact	A24	No Apparent Device Problem		
16273	2024	All other reportable incidents	47265	Dual-chamber implantable pacemaker, rate-responsive	CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A05	Mechanical Problem	22/11/2024	22/11/2024
							F26	No Health Consequences or Impact	A05	Mechanical Problem	22/11/2024	22/11/2024
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A05	Mechanical Problem	22/11/2024	22/11/2024
							F26	No Health Consequences or Impact	A05	Mechanical Problem	22/11/2024	22/11/2024
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A05	Mechanical Problem	22/11/2024	22/11/2024
							F26	No Health Consequences or Impact	A05	Mechanical Problem	22/11/2024	22/11/2024
16274	2024	All other reportable incidents	17882	Physiologic-monitoring defibrillation system	CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A25	No Apparent Adverse Event		
							F27	Problem Identified During Non-Clinical Procedure	A25	No Apparent Adverse Event		
16275	2024	All other reportable incidents	47265	Dual-chamber implantable pacemaker, rate-responsive	CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A0908	Incorrect, Inadequate or Imprecise Result or Readings	23/08/2024	23/08/2024
							F08	Hospitalization or Prolonged Hospitalization	A0908	Incorrect, Inadequate or Imprecise Result or Readings	23/08/2024	23/08/2024
							F1905	Device Revision or Replacement	A0908	Incorrect, Inadequate or Imprecise Result or Readings	23/08/2024	23/08/2024
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A0908	Incorrect, Inadequate or Imprecise Result or Readings	23/08/2024	23/08/2024
							F08	Hospitalization or Prolonged Hospitalization	A0908	Incorrect, Inadequate or Imprecise Result or Readings	23/08/2024	23/08/2024
							F1905	Device Revision or Replacement	A0908	Incorrect, Inadequate or Imprecise Result or Readings	23/08/2024	23/08/2024
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A0908	Incorrect, Inadequate or Imprecise Result or Readings	23/08/2024	23/08/2024
							F08	Hospitalization or Prolonged Hospitalization	A0908	Incorrect, Inadequate or Imprecise Result or Readings	23/08/2024	23/08/2024
							F1905	Device Revision or Replacement	A0908	Incorrect, Inadequate or Imprecise Result or Readings	23/08/2024	23/08/2024
16276	2024	All other reportable incidents	37265	Dual-chamber implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E0601	Arrhythmia	A24	No Apparent Device Problem	06/11/2024	06/11/2024
							E1901	Bacterial Infection	A24	No Apparent Device Problem	06/11/2024	06/11/2024
							F02	Death	A24	No Apparent Device Problem	06/11/2024	06/11/2024
16277	2024	All other reportable incidents			CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070909	Oversensing		

							F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A070909	Oversensing		
16278	2024	All other reportable incidents	47270	Cardiac resynchronization therapy implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E210401	Shock from Patient Lead(s)	A070909	Oversensing		
									A070910	Undersensing		
									A071302	Inappropriate/Inadequate Shock/Stimulation		
							F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A070909	Oversensing		
									A070910	Undersensing		
									A071302	Inappropriate/Inadequate Shock/Stimulation		
							F2203	Imaging Required	A070909	Oversensing		
									A070910	Undersensing		
									A071302	Inappropriate/Inadequate Shock/Stimulation		
16279	2024	All other reportable incidents	47270	Cardiac resynchronization therapy implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A040101	Fracture	15/11/2024	15/11/2024
							F08	Hospitalization or Prolonged Hospitalization	A040101	Fracture	15/11/2024	15/11/2024
							F1905	Device Revision or Replacement	A040101	Fracture	15/11/2024	15/11/2024
16280	2024	All other reportable incidents	63188	Left-ventricular wireless pacing system sensor/transmitter	CT145	Cardiovascular devices			A0708	Power Problem		
					CT1577	Wireless monitoring system transmitters			A0708	Power Problem		
					CT1684	Cardiac pacemakers and associated devices			A0708	Power Problem		
					CT2727	Active implant devices/ancillaries			A0708	Power Problem		
					CT2750	Implantable cardiac electrical stimulators and associated devices			A0708	Power Problem		
					CT2885	Physiological monitoring/measuring system transmitters/receivers/transceivers			A0708	Power Problem		
16281	2024	All other reportable incidents	47265	Dual-chamber implantable pacemaker, rate-responsive	CT1685	Cardiac pacemakers	E0601	Arrhythmia	A070909	Oversensing		
									A072001	Electromagnetic Interference		
									A1605	Reset Problem		
									A160503	Inappropriate or Unexpected Reset		
							F12	Serious Injury/ Illness/ Impairment	A070909	Oversensing		
									A072001	Electromagnetic Interference		
									A1605	Reset Problem		
									A160503	Inappropriate or Unexpected Reset		
							F2307	Reprogramming of Device	A070909	Oversensing		
									A072001	Electromagnetic Interference		
									A1605	Reset Problem		
									A160503	Inappropriate or Unexpected Reset		
					CT257	Implantable cardiac pacemakers	E0601	Arrhythmia	A070909	Oversensing		
									A072001	Electromagnetic Interference		
									A1605	Reset Problem		
									A160503	Inappropriate or Unexpected Reset		
							F12	Serious Injury/ Illness/ Impairment	A070909	Oversensing		
									A072001	Electromagnetic Interference		
									A1605	Reset Problem		
									A160503	Inappropriate or Unexpected Reset		
							F2307	Reprogramming of Device	A070909	Oversensing		
									A072001	Electromagnetic Interference		
									A1605	Reset Problem		
									A160503	Inappropriate or Unexpected Reset		
					CT2866	Implantable cardiac electrical stimulators	E0601	Arrhythmia	A070909	Oversensing		
									A072001	Electromagnetic Interference		
									A1605	Reset Problem		
									A160503	Inappropriate or Unexpected Reset		
							F12	Serious Injury/ Illness/ Impairment	A070909	Oversensing		
									A072001	Electromagnetic Interference		
									A1605	Reset Problem		
									A160503	Inappropriate or Unexpected Reset		
							F2307	Reprogramming of Device	A070909	Oversensing		
									A072001	Electromagnetic Interference		
									A1605	Reset Problem		
									A160503	Inappropriate or Unexpected Reset		
16282	2024	All other reportable incidents	17882	Physiologic-monitoring defibrillation system	CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A07	Electrical /Electronic Property Problem		
							F27	Problem Identified During Non-Clinical Procedure	A07	Electrical /Electronic Property Problem		
16283	2024	All other reportable incidents	47265	Dual-chamber implantable pacemaker, rate-responsive	CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A05	Mechanical Problem	28/11/2024	28/11/2024
							F26	No Health Consequences or Impact	A05	Mechanical Problem	28/11/2024	28/11/2024
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A05	Mechanical Problem	28/11/2024	28/11/2024
							F26	No Health Consequences or Impact	A05	Mechanical Problem	28/11/2024	28/11/2024
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A05	Mechanical Problem	28/11/2024	28/11/2024
							F26	No Health Consequences or Impact	A05	Mechanical Problem	28/11/2024	28/11/2024
16284	2024	All other reportable incidents	47270	Cardiac resynchronization therapy implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A0712	Pacing Problem		
									A071302	Inappropriate/Inadequate Shock/Stimulation		
							F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A0712	Pacing Problem		
									A071302	Inappropriate/Inadequate Shock/Stimulation		
							F2303	Medication Required	A0712	Pacing Problem		
									A071302	Inappropriate/Inadequate Shock/Stimulation		
16285	2024	All other reportable incidents	47267	Single-chamber implantable pacemaker, rate-responsive	CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A070101	Failure to Capture		
									A070102	High Capture Threshold		
									A070909	Oversensing		
									A072201	High Impedance		
									A090801	Signal Artifact/Noise		
									A1301	Failure to Read Input Signal		
							F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A070101	Failure to Capture		
									A070102	High Capture Threshold		
									A070909	Oversensing		
									A072201	High Impedance		
									A090801	Signal Artifact/Noise		
									A1301	Failure to Read Input Signal		
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A070101	Failure to Capture		
									A070102	High Capture Threshold		
									A070909	Oversensing		
									A072201	High Impedance		
									A090801	Signal Artifact/Noise		

									A1301	Failure to Read Input Signal		
							F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A070101	Failure to Capture		
									A070102	High Capture Threshold		
									A070909	Oversensing		
									A072201	High Impedance		
									A090801	Signal Artifact/Noise		
									A1301	Failure to Read Input Signal		
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A070101	Failure to Capture		
									A070102	High Capture Threshold		
									A070909	Oversensing		
									A072201	High Impedance		
									A090801	Signal Artifact/Noise		
									A1301	Failure to Read Input Signal		
							F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A070101	Failure to Capture		
									A070102	High Capture Threshold		
									A070909	Oversensing		
									A072201	High Impedance		
									A090801	Signal Artifact/Noise		
									A1301	Failure to Read Input Signal		
16286	2024	All other reportable incidents	17882	Physiologic-monitoring defibrillation system	CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A0907	No Device Output		
							F27	Problem Identified During Non-Clinical Procedure	A0907	No Device Output		
16287	2024	All other reportable incidents	17882	Physiologic-monitoring defibrillation system	CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A0908	Incorrect, Inadequate or Imprecise Result or Readings		
							F27	Problem Identified During Non-Clinical Procedure	A0908	Incorrect, Inadequate or Imprecise Result or Readings		
16288	2024	All other reportable incidents	17882	Physiologic-monitoring defibrillation system	CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A090201	Device Displays Incorrect Message		
							F26	No Health Consequences or Impact	A090201	Device Displays Incorrect Message		
16289	2024	All other reportable incidents	47270	Cardiac resynchronization therapy implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	01/09/2024	01/09/2024
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	01/09/2024	01/09/2024
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	01/09/2024	01/09/2024
16290	2024	All other reportable incidents	60789	Intracardiac pacemaker	CT1685	Cardiac pacemakers	E010701	Confusion/ Disorientation	A051201	Device Dislodged or Dislocated	29/11/2024	29/11/2024
									A07	Electrical /Electronic Property Problem	29/11/2024	29/11/2024
									A070102	High Capture Threshold	29/11/2024	29/11/2024
									A1502	Positioning Problem	29/11/2024	29/11/2024
							E0619	Pericardial Effusion	A051201	Device Dislodged or Dislocated	29/11/2024	29/11/2024
									A07	Electrical /Electronic Property Problem	29/11/2024	29/11/2024
									A070102	High Capture Threshold	29/11/2024	29/11/2024
									A1502	Positioning Problem	29/11/2024	29/11/2024
							F12	Serious Injury/ Illness/ Impairment	A051201	Device Dislodged or Dislocated	29/11/2024	29/11/2024
									A07	Electrical /Electronic Property Problem	29/11/2024	29/11/2024
									A070102	High Capture Threshold	29/11/2024	29/11/2024
									A1502	Positioning Problem	29/11/2024	29/11/2024
					CT257	Implantable cardiac pacemakers	E010701	Confusion/ Disorientation	A051201	Device Dislodged or Dislocated	29/11/2024	29/11/2024
									A07	Electrical /Electronic Property Problem	29/11/2024	29/11/2024
									A070102	High Capture Threshold	29/11/2024	29/11/2024
									A1502	Positioning Problem	29/11/2024	29/11/2024
							E0619	Pericardial Effusion	A051201	Device Dislodged or Dislocated	29/11/2024	29/11/2024
									A07	Electrical /Electronic Property Problem	29/11/2024	29/11/2024
									A070102	High Capture Threshold	29/11/2024	29/11/2024
									A1502	Positioning Problem	29/11/2024	29/11/2024
							F12	Serious Injury/ Illness/ Impairment	A051201	Device Dislodged or Dislocated	29/11/2024	29/11/2024
									A07	Electrical /Electronic Property Problem	29/11/2024	29/11/2024
									A070102	High Capture Threshold	29/11/2024	29/11/2024
									A1502	Positioning Problem	29/11/2024	29/11/2024
					CT2818	Catheterization/Cannulation/Fluid path devices	E010701	Confusion/ Disorientation	A051201	Device Dislodged or Dislocated	29/11/2024	29/11/2024
									A07	Electrical /Electronic Property Problem	29/11/2024	29/11/2024
									A070102	High Capture Threshold	29/11/2024	29/11/2024
									A1502	Positioning Problem	29/11/2024	29/11/2024
							E0619	Pericardial Effusion	A051201	Device Dislodged or Dislocated	29/11/2024	29/11/2024
									A07	Electrical /Electronic Property Problem	29/11/2024	29/11/2024
									A070102	High Capture Threshold	29/11/2024	29/11/2024
									A1502	Positioning Problem	29/11/2024	29/11/2024
							F12	Serious Injury/ Illness/ Impairment	A051201	Device Dislodged or Dislocated	29/11/2024	29/11/2024
									A07	Electrical /Electronic Property Problem	29/11/2024	29/11/2024
									A070102	High Capture Threshold	29/11/2024	29/11/2024
									A1502	Positioning Problem	29/11/2024	29/11/2024
					CT2866	Implantable cardiac electrical stimulators	E010701	Confusion/ Disorientation	A051201	Device Dislodged or Dislocated	29/11/2024	29/11/2024
									A07	Electrical /Electronic Property Problem	29/11/2024	29/11/2024
									A070102	High Capture Threshold	29/11/2024	29/11/2024
									A1502	Positioning Problem	29/11/2024	29/11/2024
							E0619	Pericardial Effusion	A051201	Device Dislodged or Dislocated	29/11/2024	29/11/2024
									A07	Electrical /Electronic Property Problem	29/11/2024	29/11/2024
									A070102	High Capture Threshold	29/11/2024	29/11/2024
									A1502	Positioning Problem	29/11/2024	29/11/2024
							F12	Serious Injury/ Illness/ Impairment	A051201	Device Dislodged or Dislocated	29/11/2024	29/11/2024
									A07	Electrical /Electronic Property Problem	29/11/2024	29/11/2024
									A070102	High Capture Threshold	29/11/2024	29/11/2024
									A1502	Positioning Problem	29/11/2024	29/11/2024
16291	2024	All other reportable incidents	17882	Physiologic-monitoring defibrillation system	CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070802	Intermittent Loss of Power		
							F26	No Health Consequences or Impact	A070802	Intermittent Loss of Power		
16292	2024	All other reportable incidents	47205	Cardiac pulse generator programmer	CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A0902	Display or Visual Feedback Problem		
									A090206	No Display/Image		
									A2202	Device Difficult to Program or Calibrate		
							F26	No Health Consequences or Impact	A0902	Display or Visual Feedback Problem		
									A090206	No Display/Image		
									A2202	Device Difficult to Program or Calibrate		
16293	2024	All other reportable incidents	47265	Dual-chamber implantable pacemaker, rate-responsive			E2403	No Clinical Signs, Symptoms or Conditions	A0908	Incorrect, Inadequate or Imprecise Result or Readings	26/11/2024	26/11/2024

							F08	Hospitalization or Prolonged Hospitalization	A0908	Incorrect, Inadequate or Imprecise Result or Readings	26/11/2024	26/11/2024
							F1905	Device Revision or Replacement	A0908	Incorrect, Inadequate or Imprecise Result or Readings	26/11/2024	26/11/2024
16294	2024	All other reportable incidents	17882	Physiologic-monitoring defibrillation system	CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A090201	Device Displays Incorrect Message		
							F27	Problem Identified During Non-Clinical Procedure	A090201	Device Displays Incorrect Message		
16295	2024	All other reportable incidents	47205	Cardiac pulse generator programmer	CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A0902	Display or Visual Feedback Problem		
							F27	Problem Identified During Non-Clinical Procedure	A0902	Display or Visual Feedback Problem		
16296	2024	All other reportable incidents	17882	Physiologic-monitoring defibrillation system	CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem		
							F27	Problem Identified During Non-Clinical Procedure	A0705	Battery Problem		
16297	2024	All other reportable incidents	36106	Implantable pacemaker analyser, intraoperative	CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A11	Computer Software Problem		
									A13	Communication or Transmission Problem		
							F26	No Health Consequences or Impact	A11	Computer Software Problem		
									A13	Communication or Transmission Problem		
16298	2024	Unanticipated serious deterioration in state of health	60789	Intracardiac pacemaker	CT1685	Cardiac pacemakers	E0619	Pericardial Effusion	A27	Appropriate Device Problem Term/Code not Available		
							E2321	Low Blood Pressure/ Hypotension	A27	Appropriate Device Problem Term/Code not Available		
							F08	Hospitalization or Prolonged Hospitalization	A27	Appropriate Device Problem Term/Code not Available		
							F12	Serious Injury/ Illness/ Impairment	A27	Appropriate Device Problem Term/Code not Available		
							F1901	Additional Surgery	A27	Appropriate Device Problem Term/Code not Available		
					CT257	Implantable cardiac pacemakers	E0619	Pericardial Effusion	A27	Appropriate Device Problem Term/Code not Available		
							E2321	Low Blood Pressure/ Hypotension	A27	Appropriate Device Problem Term/Code not Available		
							F08	Hospitalization or Prolonged Hospitalization	A27	Appropriate Device Problem Term/Code not Available		
							F12	Serious Injury/ Illness/ Impairment	A27	Appropriate Device Problem Term/Code not Available		
							F1901	Additional Surgery	A27	Appropriate Device Problem Term/Code not Available		
					CT2818	Catheterization/Cannulation/Fluid path devices	E0619	Pericardial Effusion	A27	Appropriate Device Problem Term/Code not Available		
							E2321	Low Blood Pressure/ Hypotension	A27	Appropriate Device Problem Term/Code not Available		
							F08	Hospitalization or Prolonged Hospitalization	A27	Appropriate Device Problem Term/Code not Available		
							F12	Serious Injury/ Illness/ Impairment	A27	Appropriate Device Problem Term/Code not Available		
							F1901	Additional Surgery	A27	Appropriate Device Problem Term/Code not Available		
					CT2866	Implantable cardiac electrical stimulators	E0619	Pericardial Effusion	A27	Appropriate Device Problem Term/Code not Available		
							E2321	Low Blood Pressure/ Hypotension	A27	Appropriate Device Problem Term/Code not Available		
							F08	Hospitalization or Prolonged Hospitalization	A27	Appropriate Device Problem Term/Code not Available		
							F12	Serious Injury/ Illness/ Impairment	A27	Appropriate Device Problem Term/Code not Available		
							F1901	Additional Surgery	A27	Appropriate Device Problem Term/Code not Available		
16299	2024	All other reportable incidents	47267	Single-chamber implantable pacemaker, rate-responsive	CT1685	Cardiac pacemakers	E1906	Unspecified Infection	A24	No Apparent Device Problem		
							F1203	Life Threatening Illness or Injury	A24	No Apparent Device Problem		
							F1905	Device Revision or Replacement	A24	No Apparent Device Problem		
					CT257	Implantable cardiac pacemakers	E1906	Unspecified Infection	A24	No Apparent Device Problem		
							F1203	Life Threatening Illness or Injury	A24	No Apparent Device Problem		
							F1905	Device Revision or Replacement	A24	No Apparent Device Problem		
					CT2866	Implantable cardiac electrical stimulators	E1906	Unspecified Infection	A24	No Apparent Device Problem		
							F1203	Life Threatening Illness or Injury	A24	No Apparent Device Problem		
							F1905	Device Revision or Replacement	A24	No Apparent Device Problem		
16300	2024	All other reportable incidents	35224	External pacemaker, epicardial pacing	CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A0719	Unexpected Shutdown		
									A190501	Moisture Damage		
							F23	Unexpected Medical Intervention	A0719	Unexpected Shutdown		
									A190501	Moisture Damage		
16301	2024	All other reportable incidents	17882	Physiologic-monitoring defibrillation system	CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A0719	Unexpected Shutdown		
							F27	Problem Identified During Non-Clinical Procedure	A0719	Unexpected Shutdown		
16302	2024		47265	Dual-chamber implantable pacemaker, rate-responsive	CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	13/11/2024	13/11/2024
							F1901	Additional Surgery	A24	No Apparent Device Problem	13/11/2024	13/11/2024
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	13/11/2024	13/11/2024
							F1901	Additional Surgery	A24	No Apparent Device Problem	13/11/2024	13/11/2024
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	13/11/2024	13/11/2024
							F1901	Additional Surgery	A24	No Apparent Device Problem	13/11/2024	13/11/2024
16303	2024	All other reportable incidents	47265	Dual-chamber implantable pacemaker, rate-responsive	CT1685	Cardiac pacemakers	E0601	Arrhythmia	A070909	Oversensing	27/11/2024	27/11/2024
									A070910	Undersensing	27/11/2024	27/11/2024
									A071202	Inaccurate Synchronization	27/11/2024	27/11/2024
							F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A070909	Oversensing	27/11/2024	27/11/2024
									A070910	Undersensing	27/11/2024	27/11/2024
									A071202	Inaccurate Synchronization	27/11/2024	27/11/2024
					CT257	Implantable cardiac pacemakers	E0601	Arrhythmia	A070909	Oversensing	27/11/2024	27/11/2024
									A070910	Undersensing	27/11/2024	27/11/2024
									A071202	Inaccurate Synchronization	27/11/2024	27/11/2024
							F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A070909	Oversensing	27/11/2024	27/11/2024
									A070910	Undersensing	27/11/2024	27/11/2024
									A071202	Inaccurate Synchronization	27/11/2024	27/11/2024
					CT2866	Implantable cardiac electrical stimulators	E0601	Arrhythmia	A070909	Oversensing	27/11/2024	27/11/2024
									A070910	Undersensing	27/11/2024	27/11/2024
									A071202	Inaccurate Synchronization	27/11/2024	27/11/2024
							F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A070909	Oversensing	27/11/2024	27/11/2024
									A070910	Undersensing	27/11/2024	27/11/2024
									A071202	Inaccurate Synchronization	27/11/2024	27/11/2024

16304	2024	All other reportable incidents	47270	Cardiac resynchronization therapy implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A072201	High Impedance		
							F26	No Health Consequences or Impact	A072201	High Impedance		
16305	2024	All other reportable incidents	17882	Physiologic-monitoring defibrillation system	CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A0719	Unexpected Shutdown		
							F27	Problem Identified During Non-Clinical Procedure	A0719	Unexpected Shutdown		
16306	2024	All other reportable incidents	47265	Dual-chamber implantable pacemaker, rate-responsive	CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	03/03/2025	03/03/2025
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	03/03/2025	03/03/2025
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	03/03/2025	03/03/2025
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	03/03/2025	03/03/2025
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	03/03/2025	03/03/2025
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	03/03/2025	03/03/2025
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	03/03/2025	03/03/2025
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	03/03/2025	03/03/2025
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	03/03/2025	03/03/2025
16307	2024	All other reportable incidents	35852	Single-chamber implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A0508	Noise, Audible	22/09/2024	22/09/2024
									A070504	Premature Discharge of Battery	22/09/2024	22/09/2024
									A13	Communication or Transmission Problem	22/09/2024	22/09/2024
							F08	Hospitalization or Prolonged Hospitalization	A0508	Noise, Audible	22/09/2024	22/09/2024
									A070504	Premature Discharge of Battery	22/09/2024	22/09/2024
									A13	Communication or Transmission Problem	22/09/2024	22/09/2024
							F1905	Device Revision or Replacement	A0508	Noise, Audible	22/09/2024	22/09/2024
									A070504	Premature Discharge of Battery	22/09/2024	22/09/2024
									A13	Communication or Transmission Problem	22/09/2024	22/09/2024
16308	2024	All other reportable incidents	35852	Single-chamber implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
									A0908	Incorrect, Inadequate or Imprecise Result or Readings		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
									A0908	Incorrect, Inadequate or Imprecise Result or Readings		
16309	2024	All other reportable incidents	35852	Single-chamber implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E060109	Tachycardia	A0709	Device Sensing Problem	29/11/2024	29/11/2024
									A071301	Failure to Deliver Shock/Stimulation	29/11/2024	29/11/2024
							F02	Death	A0709	Device Sensing Problem	29/11/2024	29/11/2024
									A071301	Failure to Deliver Shock/Stimulation	29/11/2024	29/11/2024
							F08	Hospitalization or Prolonged Hospitalization	A0709	Device Sensing Problem	29/11/2024	29/11/2024
									A071301	Failure to Deliver Shock/Stimulation	29/11/2024	29/11/2024
							F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A0709	Device Sensing Problem	29/11/2024	29/11/2024
									A071301	Failure to Deliver Shock/Stimulation	29/11/2024	29/11/2024
							F1203	Life Threatening Illness or Injury	A0709	Device Sensing Problem	29/11/2024	29/11/2024
									A071301	Failure to Deliver Shock/Stimulation	29/11/2024	29/11/2024
16310	2024	All other reportable incidents	35852	Single-chamber implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	29/11/2024	29/11/2024
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	29/11/2024	29/11/2024
							F12	Serious Injury/ Illness/ Impairment	A070504	Premature Discharge of Battery	29/11/2024	29/11/2024
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	29/11/2024	29/11/2024
16311	2024	All other reportable incidents	47265	Dual-chamber implantable pacemaker, rate-responsive	CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A0908	Incorrect, Inadequate or Imprecise Result or Readings	14/03/2024	14/03/2024
							F08	Hospitalization or Prolonged Hospitalization	A0908	Incorrect, Inadequate or Imprecise Result or Readings	14/03/2024	14/03/2024
							F1905	Device Revision or Replacement	A0908	Incorrect, Inadequate or Imprecise Result or Readings	14/03/2024	14/03/2024
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A0908	Incorrect, Inadequate or Imprecise Result or Readings	14/03/2024	14/03/2024
							F08	Hospitalization or Prolonged Hospitalization	A0908	Incorrect, Inadequate or Imprecise Result or Readings	14/03/2024	14/03/2024
							F1905	Device Revision or Replacement	A0908	Incorrect, Inadequate or Imprecise Result or Readings	14/03/2024	14/03/2024
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A0908	Incorrect, Inadequate or Imprecise Result or Readings	14/03/2024	14/03/2024
							F08	Hospitalization or Prolonged Hospitalization	A0908	Incorrect, Inadequate or Imprecise Result or Readings	14/03/2024	14/03/2024
							F1905	Device Revision or Replacement	A0908	Incorrect, Inadequate or Imprecise Result or Readings	14/03/2024	14/03/2024
16312	2024	All other reportable incidents	47263	Cardiac resynchronization therapy implantable pacemaker	CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A0908	Incorrect, Inadequate or Imprecise Result or Readings	03/12/2024	03/12/2024
									A1304	Telemetry Discrepancy	03/12/2024	03/12/2024
							F08	Hospitalization or Prolonged Hospitalization	A0908	Incorrect, Inadequate or Imprecise Result or Readings	03/12/2024	03/12/2024
									A1304	Telemetry Discrepancy	03/12/2024	03/12/2024
							F1905	Device Revision or Replacement	A0908	Incorrect, Inadequate or Imprecise Result or Readings	03/12/2024	03/12/2024
									A1304	Telemetry Discrepancy	03/12/2024	03/12/2024
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A0908	Incorrect, Inadequate or Imprecise Result or Readings	03/12/2024	03/12/2024
									A1304	Telemetry Discrepancy	03/12/2024	03/12/2024
							F08	Hospitalization or Prolonged Hospitalization	A0908	Incorrect, Inadequate or Imprecise Result or Readings	03/12/2024	03/12/2024
									A1304	Telemetry Discrepancy	03/12/2024	03/12/2024
							F1905	Device Revision or Replacement	A0908	Incorrect, Inadequate or Imprecise Result or Readings	03/12/2024	03/12/2024
									A1304	Telemetry Discrepancy	03/12/2024	03/12/2024
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A0908	Incorrect, Inadequate or Imprecise Result or Readings	03/12/2024	03/12/2024
									A1304	Telemetry Discrepancy	03/12/2024	03/12/2024
							F08	Hospitalization or Prolonged Hospitalization	A0908	Incorrect, Inadequate or Imprecise Result or Readings	03/12/2024	03/12/2024
									A1304	Telemetry Discrepancy	03/12/2024	03/12/2024
							F1905	Device Revision or Replacement	A0908	Incorrect, Inadequate or Imprecise Result or Readings	03/12/2024	03/12/2024
									A1304	Telemetry Discrepancy	03/12/2024	03/12/2024
16313	2024	All other reportable incidents	47263	Cardiac resynchronization therapy implantable pacemaker	CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A0908	Incorrect, Inadequate or Imprecise Result or Readings	19/11/2024	19/11/2024
							F08	Hospitalization or Prolonged Hospitalization	A0908	Incorrect, Inadequate or Imprecise Result or Readings	19/11/2024	19/11/2024
							F1905	Device Revision or Replacement	A0908	Incorrect, Inadequate or Imprecise Result or Readings	19/11/2024	19/11/2024

					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A0908	Incorrect, Inadequate or Imprecise Result or Readings	19/11/2024	19/11/2024
							F08	Hospitalization or Prolonged Hospitalization	A0908	Incorrect, Inadequate or Imprecise Result or Readings	19/11/2024	19/11/2024
							F1905	Device Revision or Replacement	A0908	Incorrect, Inadequate or Imprecise Result or Readings	19/11/2024	19/11/2024
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A0908	Incorrect, Inadequate or Imprecise Result or Readings	19/11/2024	19/11/2024
							F08	Hospitalization or Prolonged Hospitalization	A0908	Incorrect, Inadequate or Imprecise Result or Readings	19/11/2024	19/11/2024
							F1905	Device Revision or Replacement	A0908	Incorrect, Inadequate or Imprecise Result or Readings	19/11/2024	19/11/2024
16314	2024	All other reportable incidents	47270	Cardiac resynchronization therapy implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A072001	Electromagnetic Interference		
									A090801	Signal Artifact/Noise		
							F26	No Health Consequences or Impact	A072001	Electromagnetic Interference		
									A090801	Signal Artifact/Noise		
16315	2024	All other reportable incidents	47263	Cardiac resynchronization therapy implantable pacemaker	CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A1206	Misconnection	04/12/2024	04/12/2024
							F26	No Health Consequences or Impact	A1206	Misconnection	04/12/2024	04/12/2024
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A1206	Misconnection	04/12/2024	04/12/2024
							F26	No Health Consequences or Impact	A1206	Misconnection	04/12/2024	04/12/2024
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A1206	Misconnection	04/12/2024	04/12/2024
							F26	No Health Consequences or Impact	A1206	Misconnection	04/12/2024	04/12/2024
16316	2024	All other reportable incidents	47265	Dual-chamber implantable pacemaker, rate-responsive	CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A070910	Undersensing		
							F26	No Health Consequences or Impact	A070910	Undersensing		
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A070910	Undersensing		
							F26	No Health Consequences or Impact	A070910	Undersensing		
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A070910	Undersensing		
							F26	No Health Consequences or Impact	A070910	Undersensing		
16317	2024	All other reportable incidents	47270	Cardiac resynchronization therapy implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E060104	Bradycardia	A070909	Oversensing		
									A0712	Pacing Problem		
							F1001	Absence of Treatment	A070909	Oversensing		
									A0712	Pacing Problem		
							F2203	Imaging Required	A070909	Oversensing		
									A0712	Pacing Problem		
16318	2024	All other reportable incidents	36106	Implantable pacemaker analyser, intraoperative	CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A0902	Display or Visual Feedback Problem		
									A11	Computer Software Problem		
							F26	No Health Consequences or Impact	A0902	Display or Visual Feedback Problem		
									A11	Computer Software Problem		
16319	2024	All other reportable incidents	47265	Dual-chamber implantable pacemaker, rate-responsive	CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A05	Mechanical Problem	04/12/2024	04/12/2024
							F26	No Health Consequences or Impact	A05	Mechanical Problem	04/12/2024	04/12/2024
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A05	Mechanical Problem	04/12/2024	04/12/2024
							F26	No Health Consequences or Impact	A05	Mechanical Problem	04/12/2024	04/12/2024
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A05	Mechanical Problem	04/12/2024	04/12/2024
							F26	No Health Consequences or Impact	A05	Mechanical Problem	04/12/2024	04/12/2024
16320	2024	All other reportable incidents	37265	Dual-chamber implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	15/10/2024	15/10/2024
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery	15/10/2024	15/10/2024
16321	2024	All other reportable incidents	47263	Cardiac resynchronization therapy implantable pacemaker	CT1685	Cardiac pacemakers	E012207	Twitching	A0908	Incorrect, Inadequate or Imprecise Result or Readings	01/12/2024	01/12/2024
							E2311	Discomfort	A0908	Incorrect, Inadequate or Imprecise Result or Readings	01/12/2024	01/12/2024
							E2330	Pain	A0908	Incorrect, Inadequate or Imprecise Result or Readings	01/12/2024	01/12/2024
							F08	Hospitalization or Prolonged Hospitalization	A0908	Incorrect, Inadequate or Imprecise Result or Readings	01/12/2024	01/12/2024
							F1905	Device Revision or Replacement	A0908	Incorrect, Inadequate or Imprecise Result or Readings	01/12/2024	01/12/2024
					CT257	Implantable cardiac pacemakers	E012207	Twitching	A0908	Incorrect, Inadequate or Imprecise Result or Readings	01/12/2024	01/12/2024
							E2311	Discomfort	A0908	Incorrect, Inadequate or Imprecise Result or Readings	01/12/2024	01/12/2024
							E2330	Pain	A0908	Incorrect, Inadequate or Imprecise Result or Readings	01/12/2024	01/12/2024
							F08	Hospitalization or Prolonged Hospitalization	A0908	Incorrect, Inadequate or Imprecise Result or Readings	01/12/2024	01/12/2024
							F1905	Device Revision or Replacement	A0908	Incorrect, Inadequate or Imprecise Result or Readings	01/12/2024	01/12/2024
					CT2866	Implantable cardiac electrical stimulators	E012207	Twitching	A0908	Incorrect, Inadequate or Imprecise Result or Readings	01/12/2024	01/12/2024
							E2311	Discomfort	A0908	Incorrect, Inadequate or Imprecise Result or Readings	01/12/2024	01/12/2024
							E2330	Pain	A0908	Incorrect, Inadequate or Imprecise Result or Readings	01/12/2024	01/12/2024
							F08	Hospitalization or Prolonged Hospitalization	A0908	Incorrect, Inadequate or Imprecise Result or Readings	01/12/2024	01/12/2024
							F1905	Device Revision or Replacement	A0908	Incorrect, Inadequate or Imprecise Result or Readings	01/12/2024	01/12/2024
16322	2024	All other reportable incidents	47265	Dual-chamber implantable pacemaker, rate-responsive	CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A090804	False Positive Result		
							F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A090804	False Positive Result		
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A090804	False Positive Result		
							F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A090804	False Positive Result		
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A090804	False Positive Result		
							F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A090804	False Positive Result		
16323	2024	All other reportable incidents	47270	Cardiac resynchronization therapy implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E210401	Shock from Patient Lead(s)	A070909	Oversensing		
									A071302	Inappropriate/Inadequate Shock/Stimulation		
							F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A070909	Oversensing		
									A071302	Inappropriate/Inadequate Shock/Stimulation		
16324	2024	All other reportable incidents	47265	Dual-chamber implantable pacemaker, rate-responsive	CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	18/08/2024	18/08/2024
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	18/08/2024	18/08/2024
							F1903	Device Explantation	A070504	Premature Discharge of Battery	18/08/2024	18/08/2024
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	18/08/2024	18/08/2024
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	18/08/2024	18/08/2024
							F1903	Device Explantation	A070504	Premature Discharge of Battery	18/08/2024	18/08/2024
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	18/08/2024	18/08/2024

							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	18/08/2024	18/08/2024
							F1903	Device Explantation	A070504	Premature Discharge of Battery	18/08/2024	18/08/2024
16325	2024	All other reportable incidents			CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070909	Oversensing		
									A0712	Pacing Problem		
									A072201	High Impedance		
							F1001	Absence of Treatment	A070909	Oversensing		
									A0712	Pacing Problem		
									A072201	High Impedance		
16326	2024	All other reportable incidents	47263	Cardiac resynchronization therapy implantable pacemaker	CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A0908	Incorrect, Inadequate or Imprecise Result or Readings	03/09/2024	03/09/2024
							F08	Hospitalization or Prolonged Hospitalization	A0908	Incorrect, Inadequate or Imprecise Result or Readings	03/09/2024	03/09/2024
							F1905	Device Revision or Replacement	A0908	Incorrect, Inadequate or Imprecise Result or Readings	03/09/2024	03/09/2024
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A0908	Incorrect, Inadequate or Imprecise Result or Readings	03/09/2024	03/09/2024
							F08	Hospitalization or Prolonged Hospitalization	A0908	Incorrect, Inadequate or Imprecise Result or Readings	03/09/2024	03/09/2024
							F1905	Device Revision or Replacement	A0908	Incorrect, Inadequate or Imprecise Result or Readings	03/09/2024	03/09/2024
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A0908	Incorrect, Inadequate or Imprecise Result or Readings	03/09/2024	03/09/2024
							F08	Hospitalization or Prolonged Hospitalization	A0908	Incorrect, Inadequate or Imprecise Result or Readings	03/09/2024	03/09/2024
							F1905	Device Revision or Replacement	A0908	Incorrect, Inadequate or Imprecise Result or Readings	03/09/2024	03/09/2024
16327	2024		47263	Cardiac resynchronization therapy implantable pacemaker	CT1685	Cardiac pacemakers	E0119	Loss of Consciousness	A01	Patient Device Interaction Problem		
									A070909	Oversensing		
									A11	Computer Software Problem		
							F08	Hospitalization or Prolonged Hospitalization	A01	Patient Device Interaction Problem		
									A070909	Oversensing		
									A11	Computer Software Problem		
							F11	Non-Serious Injury/ Illness/ Impairment	A01	Patient Device Interaction Problem		
									A070909	Oversensing		
									A11	Computer Software Problem		
					CT257	Implantable cardiac pacemakers	E0119	Loss of Consciousness	A01	Patient Device Interaction Problem		
									A070909	Oversensing		
									A11	Computer Software Problem		
							F08	Hospitalization or Prolonged Hospitalization	A01	Patient Device Interaction Problem		
									A070909	Oversensing		
									A11	Computer Software Problem		
							F11	Non-Serious Injury/ Illness/ Impairment	A01	Patient Device Interaction Problem		
									A070909	Oversensing		
									A11	Computer Software Problem		
					CT2866	Implantable cardiac electrical stimulators	E0119	Loss of Consciousness	A01	Patient Device Interaction Problem		
									A070909	Oversensing		
									A11	Computer Software Problem		
							F08	Hospitalization or Prolonged Hospitalization	A01	Patient Device Interaction Problem		
									A070909	Oversensing		
									A11	Computer Software Problem		
							F11	Non-Serious Injury/ Illness/ Impairment	A01	Patient Device Interaction Problem		
									A070909	Oversensing		
									A11	Computer Software Problem		
16328	2024		35852	Single-chamber implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery		
16329	2024	All other reportable incidents	17882	Physiologic-monitoring defibrillation system	CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A0719	Unexpected Shutdown		
							F27	Problem Identified During Non-Clinical Procedure	A0719	Unexpected Shutdown		
16330	2024	All other reportable incidents	37265	Dual-chamber implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E060104	Bradycardia	A0712	Pacing Problem		
							F12	Serious Injury/ Illness/ Impairment	A0712	Pacing Problem		
16331	2024	All other reportable incidents	35852	Single-chamber implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A070912	Incorrect Interpretation of Signal		
									A071302	Inappropriate/Inadequate Shock/Stimulation		
							F12	Serious Injury/ Illness/ Impairment	A070912	Incorrect Interpretation of Signal		
									A071302	Inappropriate/Inadequate Shock/Stimulation		
16332	2024	All other reportable incidents	47267	Single-chamber implantable pacemaker, rate-responsive	CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A070909	Oversensing		
							F22	Unexpected Diagnostic Intervention	A070909	Oversensing		
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A070909	Oversensing		
							F22	Unexpected Diagnostic Intervention	A070909	Oversensing		
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A070909	Oversensing		
							F22	Unexpected Diagnostic Intervention	A070909	Oversensing		
16333	2024	All other reportable incidents	47270	Cardiac resynchronization therapy implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
16334	2024	All other reportable incidents	17882	Physiologic-monitoring defibrillation system	CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A1603	Failure of Device to Self-Test		
							F27	Problem Identified During Non-Clinical Procedure	A1603	Failure of Device to Self-Test		
16335	2024	All other reportable incidents	47270	Cardiac resynchronization therapy implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem	28/01/2025	28/01/2025
									A070504	Premature Discharge of Battery	28/01/2025	28/01/2025
							F08	Hospitalization or Prolonged Hospitalization	A0705	Battery Problem	28/01/2025	28/01/2025
									A070504	Premature Discharge of Battery	28/01/2025	28/01/2025
							F1905	Device Revision or Replacement	A0705	Battery Problem	28/01/2025	28/01/2025
									A070504	Premature Discharge of Battery	28/01/2025	28/01/2025
16336	2024	Unanticipated serious deterioration in state of health	47265	Dual-chamber implantable pacemaker, rate-responsive	CT1685	Cardiac pacemakers	E011903	Syncope/Fainting	A0712	Pacing Problem	04/12/2024	04/12/2024
							E0602	Cardiac Arrest	A0712	Pacing Problem	04/12/2024	04/12/2024
							F08	Hospitalization or Prolonged Hospitalization	A0712	Pacing Problem	04/12/2024	04/12/2024
							F1203	Life Threatening Illness or Injury	A0712	Pacing Problem	04/12/2024	04/12/2024
							F1905	Device Revision or Replacement	A0712	Pacing Problem	04/12/2024	04/12/2024
					CT257	Implantable cardiac pacemakers	E011903	Syncope/Fainting	A0712	Pacing Problem	04/12/2024	04/12/2024
							E0602	Cardiac Arrest	A0712	Pacing Problem	04/12/2024	04/12/2024
							F08	Hospitalization or Prolonged Hospitalization	A0712	Pacing Problem	04/12/2024	04/12/2024
							F1203	Life Threatening Illness or Injury	A0712	Pacing Problem	04/12/2024	04/12/2024

							F1905	Device Revision or Replacement	A0712	Pacing Problem	04/12/2024	04/12/2024
					CT2866	Implantable cardiac electrical stimulators	E011903	Syncope/Fainting	A0712	Pacing Problem	04/12/2024	04/12/2024
							E0602	Cardiac Arrest	A0712	Pacing Problem	04/12/2024	04/12/2024
							F08	Hospitalization or Prolonged Hospitalization	A0712	Pacing Problem	04/12/2024	04/12/2024
							F1203	Life Threatening Illness or Injury	A0712	Pacing Problem	04/12/2024	04/12/2024
							F1905	Device Revision or Replacement	A0712	Pacing Problem	04/12/2024	04/12/2024
16337	2024	All other reportable incidents	47270	Cardiac resynchronization therapy implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E060109	Tachycardia	A0709	Device Sensing Problem		
									A0713	Defibrillation/Stimulation Problem		
							F05	Delay to Treatment/ Therapy	A0709	Device Sensing Problem		
									A0713	Defibrillation/Stimulation Problem		
							F1203	Life Threatening Illness or Injury	A0709	Device Sensing Problem		
									A0713	Defibrillation/Stimulation Problem		
							F2307	Reprogramming of Device	A0709	Device Sensing Problem		
									A0713	Defibrillation/Stimulation Problem		
16338	2024	All other reportable incidents	47265	Dual-chamber implantable pacemaker, rate-responsive	CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	06/12/2024	06/12/2024
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	06/12/2024	06/12/2024
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	06/12/2024	06/12/2024
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	06/12/2024	06/12/2024
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	06/12/2024	06/12/2024
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	06/12/2024	06/12/2024
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	06/12/2024	06/12/2024
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	06/12/2024	06/12/2024
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	06/12/2024	06/12/2024
16339	2024	All other reportable incidents	47270	Cardiac resynchronization therapy implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	01/03/2025	01/03/2025
									A0722	Impedance Problem	01/03/2025	01/03/2025
									A0908	Incorrect, Inadequate or Imprecise Result or Readings	01/03/2025	01/03/2025
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	01/03/2025	01/03/2025
									A0722	Impedance Problem	01/03/2025	01/03/2025
									A0908	Incorrect, Inadequate or Imprecise Result or Readings	01/03/2025	01/03/2025
							F1903	Device Explantation	A070504	Premature Discharge of Battery	01/03/2025	01/03/2025
									A0722	Impedance Problem	01/03/2025	01/03/2025
									A0908	Incorrect, Inadequate or Imprecise Result or Readings	01/03/2025	01/03/2025
16340	2024	All other reportable incidents	47270	Cardiac resynchronization therapy implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A072202	Low Impedance		
							F2203	Imaging Required	A072202	Low Impedance		
16341	2024	All other reportable incidents	17882	Physiologic-monitoring defibrillation system	CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A1603	Failure of Device to Self-Test		
							F27	Problem Identified During Non-Clinical Procedure	A1603	Failure of Device to Self-Test		
16342	2024	All other reportable incidents	17882	Physiologic-monitoring defibrillation system	CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A07	Electrical /Electronic Property Problem		
							F27	Problem Identified During Non-Clinical Procedure	A07	Electrical /Electronic Property Problem		
16343	2024	All other reportable incidents	47265	Dual-chamber implantable pacemaker, rate-responsive	CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
									A1605	Reset Problem		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
									A1605	Reset Problem		
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
									A1605	Reset Problem		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
									A1605	Reset Problem		
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
									A1605	Reset Problem		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
									A1605	Reset Problem		
16344	2024	All other reportable incidents	47267	Single-chamber implantable pacemaker, rate-responsive	CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A070101	Failure to Capture		
									A070909	Oversensing		
							F26	No Health Consequences or Impact	A070101	Failure to Capture		
									A070909	Oversensing		
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A070101	Failure to Capture		
									A070909	Oversensing		
							F26	No Health Consequences or Impact	A070101	Failure to Capture		
									A070909	Oversensing		
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A070101	Failure to Capture		
									A070909	Oversensing		
							F26	No Health Consequences or Impact	A070101	Failure to Capture		
									A070909	Oversensing		
16345	2024	All other reportable incidents	47265	Dual-chamber implantable pacemaker, rate-responsive	CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
16346	2024	All other reportable incidents	60789	Intracardiac pacemaker	CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A070102	High Capture Threshold		
									A090801	Signal Artifact/Noise		
							F2307	Reprogramming of Device	A070102	High Capture Threshold		
									A090801	Signal Artifact/Noise		
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A070102	High Capture Threshold		
									A090801	Signal Artifact/Noise		
							F2307	Reprogramming of Device	A070102	High Capture Threshold		
									A090801	Signal Artifact/Noise		
					CT2818	Catheterization/Cannulation/Fluid path devices	E2403	No Clinical Signs, Symptoms or Conditions	A070102	High Capture Threshold		
									A090801	Signal Artifact/Noise		
							F2307	Reprogramming of Device	A070102	High Capture Threshold		
									A090801	Signal Artifact/Noise		
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A070102	High Capture Threshold		
									A090801	Signal Artifact/Noise		
							F2307	Reprogramming of Device	A070102	High Capture Threshold		
									A090801	Signal Artifact/Noise		
16347	2024	All other reportable incidents	37265	Dual-chamber implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A160601	Premature Elective Replacement Indicator		

							F1903	Device Explantation	A160601	Premature Elective Replacement Indicator		
16348	2024	All other reportable incidents	36106	Implantable pacemaker analyser, intraoperative	CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A0902	Display or Visual Feedback Problem		
									A11	Computer Software Problem		
							F26	No Health Consequences or Impact	A0902	Display or Visual Feedback Problem		
									A11	Computer Software Problem		
16349	2024	All other reportable incidents	17882	Physiologic-monitoring defibrillation system	CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070603	Failure to Charge		
									A071301	Failure to Deliver Shock/Stimulation		
									A0719	Unexpected Shutdown		
							F27	Problem Identified During Non-Clinical Procedure	A070603	Failure to Charge		
									A071301	Failure to Deliver Shock/Stimulation		
									A0719	Unexpected Shutdown		
16350	2024		47205	Cardiac pulse generator programmer	CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A090801	Signal Artifact/Noise		
							F26	No Health Consequences or Impact	A090801	Signal Artifact/Noise		
16351	2024	All other reportable incidents	47270	Cardiac resynchronization therapy implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A072201	High Impedance		
							F26	No Health Consequences or Impact	A072201	High Impedance		
16352	2024	All other reportable incidents	47270	Cardiac resynchronization therapy implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A010402	Migration		
									A051201	Device Dislodged or Dislocated		
							F26	No Health Consequences or Impact	A010402	Migration		
									A051201	Device Dislodged or Dislocated		
16353	2024	All other reportable incidents	47265	Dual-chamber implantable pacemaker, rate-responsive	CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A0908	Incorrect, Inadequate or Imprecise Result or Readings	04/02/2025	04/02/2025
							F08	Hospitalization or Prolonged Hospitalization	A0908	Incorrect, Inadequate or Imprecise Result or Readings	04/02/2025	04/02/2025
							F1903	Device Explantation	A0908	Incorrect, Inadequate or Imprecise Result or Readings	04/02/2025	04/02/2025
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A0908	Incorrect, Inadequate or Imprecise Result or Readings	04/02/2025	04/02/2025
							F08	Hospitalization or Prolonged Hospitalization	A0908	Incorrect, Inadequate or Imprecise Result or Readings	04/02/2025	04/02/2025
							F1903	Device Explantation	A0908	Incorrect, Inadequate or Imprecise Result or Readings	04/02/2025	04/02/2025
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A0908	Incorrect, Inadequate or Imprecise Result or Readings	04/02/2025	04/02/2025
							F08	Hospitalization or Prolonged Hospitalization	A0908	Incorrect, Inadequate or Imprecise Result or Readings	04/02/2025	04/02/2025
							F1903	Device Explantation	A0908	Incorrect, Inadequate or Imprecise Result or Readings	04/02/2025	04/02/2025
16354	2024		18111	Pacing lead extraction kit, reusable	CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A071205	Pacemaker Found in Back-Up Mode		
							F1905	Device Revision or Replacement	A071205	Pacemaker Found in Back-Up Mode		
16355	2024		47265	Dual-chamber implantable pacemaker, rate-responsive	CT1685	Cardiac pacemakers	E2401	Insufficient Information	A01	Patient Device Interaction Problem		
									A0711	Interrogation Problem		
									A160202	No Fail-Safe Mechanism		
							F01	Change in Therapeutic Response	A01	Patient Device Interaction Problem		
									A0711	Interrogation Problem		
									A160202	No Fail-Safe Mechanism		
					CT257	Implantable cardiac pacemakers	E2401	Insufficient Information	A01	Patient Device Interaction Problem		
									A0711	Interrogation Problem		
									A160202	No Fail-Safe Mechanism		
							F01	Change in Therapeutic Response	A01	Patient Device Interaction Problem		
									A0711	Interrogation Problem		
									A160202	No Fail-Safe Mechanism		
					CT2866	Implantable cardiac electrical stimulators	E2401	Insufficient Information	A01	Patient Device Interaction Problem		
									A0711	Interrogation Problem		
									A160202	No Fail-Safe Mechanism		
							F01	Change in Therapeutic Response	A01	Patient Device Interaction Problem		
									A0711	Interrogation Problem		
									A160202	No Fail-Safe Mechanism		
16356	2024	All other reportable incidents	17882	Physiologic-monitoring defibrillation system	CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A0709	Device Sensing Problem		
							F26	No Health Consequences or Impact	A0709	Device Sensing Problem		
16357	2024	All other reportable incidents	47270	Cardiac resynchronization therapy implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A05	Mechanical Problem	09/12/2024	09/12/2024
									A072201	High Impedance	09/12/2024	09/12/2024
							F26	No Health Consequences or Impact	A05	Mechanical Problem	09/12/2024	09/12/2024
									A072201	High Impedance	09/12/2024	09/12/2024
16358	2024	All other reportable incidents	17882	Physiologic-monitoring defibrillation system	CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A07	Electrical /Electronic Property Problem		
							F26	No Health Consequences or Impact	A07	Electrical /Electronic Property Problem		
16359	2024	All other reportable incidents	47265	Dual-chamber implantable pacemaker, rate-responsive	CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem		
							F1901	Additional Surgery	A24	No Apparent Device Problem		
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem		
							F1901	Additional Surgery	A24	No Apparent Device Problem		
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem		
							F1901	Additional Surgery	A24	No Apparent Device Problem		
16360	2024	All other reportable incidents	17882	Physiologic-monitoring defibrillation system	CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A07	Electrical /Electronic Property Problem		
							F27	Problem Identified During Non-Clinical Procedure	A07	Electrical /Electronic Property Problem		
16361	2024	All other reportable incidents	47270	Cardiac resynchronization therapy implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	26/03/2025	26/03/2025
									A0908	Incorrect, Inadequate or Imprecise Result or Readings	26/03/2025	26/03/2025
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	26/03/2025	26/03/2025
									A0908	Incorrect, Inadequate or Imprecise Result or Readings	26/03/2025	26/03/2025
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	26/03/2025	26/03/2025
									A0908	Incorrect, Inadequate or Imprecise Result or Readings	26/03/2025	26/03/2025
16362	2024	All other reportable incidents	47267	Single-chamber implantable pacemaker, rate-responsive	CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
16363	2024	All other reportable incidents	47265	Dual-chamber implantable pacemaker, rate-responsive	CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	06/12/2024	06/12/2024
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	06/12/2024	06/12/2024
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	06/12/2024	06/12/2024

					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	06/12/2024	06/12/2024
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	06/12/2024	06/12/2024
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	06/12/2024	06/12/2024
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	06/12/2024	06/12/2024
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	06/12/2024	06/12/2024
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	06/12/2024	06/12/2024
16364	2024	All other reportable incidents	17882	Physiologic-monitoring defibrillation system	CT1684	Cardiac pacemakers and associated devices	E1704	Burn(s)	A0704	Sparking		
							F11	Non-Serious Injury/ Illness/ Impairment	A0704	Sparking		
16365	2024	All other reportable incidents			CT1684	Cardiac pacemakers and associated devices	E0601	Arrhythmia	A070101	Failure to Capture		
							F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A070101	Failure to Capture		
							F15	Recognized Device or Procedural Complication	A070101	Failure to Capture		
16366	2024	All other reportable incidents	17882	Physiologic-monitoring defibrillation system	CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A0801	Failure to Calibrate		
							F27	Problem Identified During Non-Clinical Procedure	A0801	Failure to Calibrate		
16367	2024	All other reportable incidents	17882	Physiologic-monitoring defibrillation system	CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A0801	Failure to Calibrate		
							F27	Problem Identified During Non-Clinical Procedure	A0801	Failure to Calibrate		
16368	2024	All other reportable incidents	17882	Physiologic-monitoring defibrillation system	CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A0801	Failure to Calibrate		
							F27	Problem Identified During Non-Clinical Procedure	A0801	Failure to Calibrate		
16369	2024	All other reportable incidents	17882	Physiologic-monitoring defibrillation system	CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A0801	Failure to Calibrate		
							F27	Problem Identified During Non-Clinical Procedure	A0801	Failure to Calibrate		
16370	2024	All other reportable incidents	60789	Intracardiac pacemaker	CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A0701	Capturing Problem		
									A070102	High Capture Threshold		
							F26	No Health Consequences or Impact	A0701	Capturing Problem		
									A070102	High Capture Threshold		
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A0701	Capturing Problem		
									A070102	High Capture Threshold		
							F26	No Health Consequences or Impact	A0701	Capturing Problem		
									A070102	High Capture Threshold		
					CT2818	Catheterization/Cannulation/Fluid path devices	E2403	No Clinical Signs, Symptoms or Conditions	A0701	Capturing Problem		
									A070102	High Capture Threshold		
							F26	No Health Consequences or Impact	A0701	Capturing Problem		
									A070102	High Capture Threshold		
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A0701	Capturing Problem		
									A070102	High Capture Threshold		
							F26	No Health Consequences or Impact	A0701	Capturing Problem		
									A070102	High Capture Threshold		
16371	2024	All other reportable incidents	47270	Cardiac resynchronization therapy implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A0712	Pacing Problem		
							F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A0712	Pacing Problem		
16372	2024	All other reportable incidents	17882	Physiologic-monitoring defibrillation system	CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070912	Incorrect Interpretation of Signal		
							F26	No Health Consequences or Impact	A070912	Incorrect Interpretation of Signal		
16373	2024	All other reportable incidents	47205	Cardiac pulse generator programmer	CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A0203	Defective Device		
							F26	No Health Consequences or Impact	A0203	Defective Device		
16374	2024		47267	Single-chamber implantable pacemaker, rate-responsive	CT1685	Cardiac pacemakers	E2312	Fatigue	A070504	Premature Discharge of Battery	26/11/2024	26/11/2024
									A071204	No Pacing	26/11/2024	26/11/2024
							F1901	Additional Surgery	A070504	Premature Discharge of Battery	26/11/2024	26/11/2024
									A071204	No Pacing	26/11/2024	26/11/2024
					CT257	Implantable cardiac pacemakers	E2312	Fatigue	A070504	Premature Discharge of Battery	26/11/2024	26/11/2024
									A071204	No Pacing	26/11/2024	26/11/2024
							F1901	Additional Surgery	A070504	Premature Discharge of Battery	26/11/2024	26/11/2024
									A071204	No Pacing	26/11/2024	26/11/2024
					CT2866	Implantable cardiac electrical stimulators	E2312	Fatigue	A070504	Premature Discharge of Battery	26/11/2024	26/11/2024
									A071204	No Pacing	26/11/2024	26/11/2024
							F1901	Additional Surgery	A070504	Premature Discharge of Battery	26/11/2024	26/11/2024
									A071204	No Pacing	26/11/2024	26/11/2024
16375	2024	All other reportable incidents	60956	Pacing/defibrillation lead tunneller	CT1684	Cardiac pacemakers and associated devices	E210401	Shock from Patient Lead(s)	A070909	Oversensing		
									A071302	Inappropriate/Inadequate Shock/Stimulation		
									A090801	Signal Artifact/Noise		
									A150202	Malposition of Device		
							F08	Hospitalization or Prolonged Hospitalization	A070909	Oversensing		
									A071302	Inappropriate/Inadequate Shock/Stimulation		
									A090801	Signal Artifact/Noise		
									A150202	Malposition of Device		
							F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A070909	Oversensing		
									A071302	Inappropriate/Inadequate Shock/Stimulation		
									A090801	Signal Artifact/Noise		
									A150202	Malposition of Device		
							F2203	Imaging Required	A070909	Oversensing		
									A071302	Inappropriate/Inadequate Shock/Stimulation		
									A090801	Signal Artifact/Noise		
									A150202	Malposition of Device		
							F23	Unexpected Medical Intervention	A070909	Oversensing		
									A071302	Inappropriate/Inadequate Shock/Stimulation		
									A090801	Signal Artifact/Noise		
									A150202	Malposition of Device		
16376	2024		47265	Dual-chamber implantable pacemaker, rate-responsive	CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A1204	Failure to Disconnect	12/12/2024	12/12/2024
									A1205	Loose or Intermittent Connection	12/12/2024	12/12/2024
							F1901	Additional Surgery	A1204	Failure to Disconnect	12/12/2024	12/12/2024
									A1205	Loose or Intermittent Connection	12/12/2024	12/12/2024
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A1204	Failure to Disconnect	12/12/2024	12/12/2024
									A1205	Loose or Intermittent Connection	12/12/2024	12/12/2024
							F1901	Additional Surgery	A1204	Failure to Disconnect	12/12/2024	12/12/2024
									A1205	Loose or Intermittent Connection	12/12/2024	12/12/2024
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A1204	Failure to Disconnect	12/12/2024	12/12/2024
									A1205	Loose or Intermittent Connection	12/12/2024	12/12/2024
							F1901	Additional Surgery	A1204	Failure to Disconnect	12/12/2024	12/12/2024
									A1205	Loose or Intermittent Connection	12/12/2024	12/12/2024
16377	2024	All other reportable incidents	60789	Intracardiac pacemaker	CT1685	Cardiac pacemakers	E0601	Arrhythmia	A051201	Device Dislodged or Dislocated		
									A070101	Failure to Capture		
									A072202	Low Impedance		

							E0604	Cardiac Perforation	A051201	Device Dislodged or Dislocated		
									A070101	Failure to Capture		
									A072202	Low Impedance		
							E0605	Cardiac Tamponade	A051201	Device Dislodged or Dislocated		
									A070101	Failure to Capture		
									A072202	Low Impedance		
							E0619	Pericardial Effusion	A051201	Device Dislodged or Dislocated		
									A070101	Failure to Capture		
									A072202	Low Impedance		
							F1203	Life Threatening Illness or Injury	A051201	Device Dislodged or Dislocated		
									A070101	Failure to Capture		
									A072202	Low Impedance		
					CT257	Implantable cardiac pacemakers	E0601	Arrhythmia	A051201	Device Dislodged or Dislocated		
									A070101	Failure to Capture		
									A072202	Low Impedance		
							E0604	Cardiac Perforation	A051201	Device Dislodged or Dislocated		
									A070101	Failure to Capture		
									A072202	Low Impedance		
							E0605	Cardiac Tamponade	A051201	Device Dislodged or Dislocated		
									A070101	Failure to Capture		
									A072202	Low Impedance		
							E0619	Pericardial Effusion	A051201	Device Dislodged or Dislocated		
									A070101	Failure to Capture		
									A072202	Low Impedance		
							F1203	Life Threatening Illness or Injury	A051201	Device Dislodged or Dislocated		
									A070101	Failure to Capture		
									A072202	Low Impedance		
					CT2818	Catheterization/Cannulation/Fluid path devices	E0601	Arrhythmia	A051201	Device Dislodged or Dislocated		
									A070101	Failure to Capture		
									A072202	Low Impedance		
							E0604	Cardiac Perforation	A051201	Device Dislodged or Dislocated		
									A070101	Failure to Capture		
									A072202	Low Impedance		
							E0605	Cardiac Tamponade	A051201	Device Dislodged or Dislocated		
									A070101	Failure to Capture		
									A072202	Low Impedance		
							E0619	Pericardial Effusion	A051201	Device Dislodged or Dislocated		
									A070101	Failure to Capture		
									A072202	Low Impedance		
							F1203	Life Threatening Illness or Injury	A051201	Device Dislodged or Dislocated		
									A070101	Failure to Capture		
									A072202	Low Impedance		
					CT2866	Implantable cardiac electrical stimulators	E0601	Arrhythmia	A051201	Device Dislodged or Dislocated		
									A070101	Failure to Capture		
									A072202	Low Impedance		
							E0604	Cardiac Perforation	A051201	Device Dislodged or Dislocated		
									A070101	Failure to Capture		
									A072202	Low Impedance		
							E0605	Cardiac Tamponade	A051201	Device Dislodged or Dislocated		
									A070101	Failure to Capture		
									A072202	Low Impedance		
							E0619	Pericardial Effusion	A051201	Device Dislodged or Dislocated		
									A070101	Failure to Capture		
									A072202	Low Impedance		
							F1203	Life Threatening Illness or Injury	A051201	Device Dislodged or Dislocated		
									A070101	Failure to Capture		
									A072202	Low Impedance		
16378	2024	All other reportable incidents	57968	Cardiovascular device introducer, steerable	CT1684	Cardiac pacemakers and associated devices	E0601	Arrhythmia	A24	No Apparent Device Problem		
							E0604	Cardiac Perforation	A24	No Apparent Device Problem		
							E0605	Cardiac Tamponade	A24	No Apparent Device Problem		
							E0619	Pericardial Effusion	A24	No Apparent Device Problem		
							F1203	Life Threatening Illness or Injury	A24	No Apparent Device Problem		
16379	2024	All other reportable incidents	35852	Single-chamber implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A160601	Premature Elective Replacement Indicator	12/12/2024	12/12/2024
							F1903	Device Explantation	A160601	Premature Elective Replacement Indicator	12/12/2024	12/12/2024
16380	2024	All other reportable incidents	47265	Dual-chamber implantable pacemaker, rate-responsive	CT1685	Cardiac pacemakers	E011902	Presyncope	A0712	Pacing Problem	14/01/2025	14/01/2025
									A1304	Telemetry Discrepancy	14/01/2025	14/01/2025
							E233001	Chest Pain	A0712	Pacing Problem	14/01/2025	14/01/2025
									A1304	Telemetry Discrepancy	14/01/2025	14/01/2025
							F08	Hospitalization or Prolonged Hospitalization	A0712	Pacing Problem	14/01/2025	14/01/2025
									A1304	Telemetry Discrepancy	14/01/2025	14/01/2025
							F1001	Absence of Treatment	A0712	Pacing Problem	14/01/2025	14/01/2025
									A1304	Telemetry Discrepancy	14/01/2025	14/01/2025
							F1905	Device Revision or Replacement	A0712	Pacing Problem	14/01/2025	14/01/2025
									A1304	Telemetry Discrepancy	14/01/2025	14/01/2025
					CT257	Implantable cardiac pacemakers	E011902	Presyncope	A0712	Pacing Problem	14/01/2025	14/01/2025
									A1304	Telemetry Discrepancy	14/01/2025	14/01/2025
							E233001	Chest Pain	A0712	Pacing Problem	14/01/2025	14/01/2025
									A1304	Telemetry Discrepancy	14/01/2025	14/01/2025
							F08	Hospitalization or Prolonged Hospitalization	A0712	Pacing Problem	14/01/2025	14/01/2025
									A1304	Telemetry Discrepancy	14/01/2025	14/01/2025
							F1001	Absence of Treatment	A0712	Pacing Problem	14/01/2025	14/01/2025
									A1304	Telemetry Discrepancy	14/01/2025	14/01/2025
							F1905	Device Revision or Replacement	A0712	Pacing Problem	14/01/2025	14/01/2025
									A1304	Telemetry Discrepancy	14/01/2025	14/01/2025
					CT2866	Implantable cardiac electrical stimulators	E011902	Presyncope	A0712	Pacing Problem	14/01/2025	14/01/2025
									A1304	Telemetry Discrepancy	14/01/2025	14/01/2025
							E233001	Chest Pain	A0712	Pacing Problem	14/01/2025	14/01/2025
									A1304	Telemetry Discrepancy	14/01/2025	14/01/2025

							F08	Hospitalization or Prolonged Hospitalization	A0712	Pacing Problem	14/01/2025	14/01/2025
									A1304	Telemetry Discrepancy	14/01/2025	14/01/2025
							F1001	Absence of Treatment	A0712	Pacing Problem	14/01/2025	14/01/2025
									A1304	Telemetry Discrepancy	14/01/2025	14/01/2025
							F1905	Device Revision or Replacement	A0712	Pacing Problem	14/01/2025	14/01/2025
									A1304	Telemetry Discrepancy	14/01/2025	14/01/2025
16381	2024	All other reportable incidents	35224	External pacemaker, epicardial pacing	CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A07	Electrical /Electronic Property Problem		
							F27	Problem Identified During Non-Clinical Procedure	A07	Electrical /Electronic Property Problem		
16382	2024	All other reportable incidents	47270	Cardiac resynchronization therapy implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E0138	Undesired Nerve Stimulation	A0101	Patient-Device Incompatibility		
									A070909	Oversensing		
									A090801	Signal Artifact/Noise		
							F22	Unexpected Diagnostic Intervention	A0101	Patient-Device Incompatibility		
									A070909	Oversensing		
									A090801	Signal Artifact/Noise		
							F2306	Resuscitation	A0101	Patient-Device Incompatibility		
									A070909	Oversensing		
									A090801	Signal Artifact/Noise		
16383	2024	All other reportable incidents	47205	Cardiac pulse generator programmer	CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A09	Output Problem		
									A0902	Display or Visual Feedback Problem		
									A090801	Signal Artifact/Noise		
									A13	Communication or Transmission Problem		
							F23	Unexpected Medical Intervention	A09	Output Problem		
									A0902	Display or Visual Feedback Problem		
									A090801	Signal Artifact/Noise		
									A13	Communication or Transmission Problem		
16384	2024	All other reportable incidents	17882	Physiologic-monitoring defibrillation system	CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070803	Failure to Power Up		
							F27	Problem Identified During Non-Clinical Procedure	A070803	Failure to Power Up		
16385	2024	All other reportable incidents	47263	Cardiac resynchronization therapy implantable pacemaker	CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A070909	Oversensing		
									A090801	Signal Artifact/Noise		
									A1702	Device-Device Incompatibility		
							F26	No Health Consequences or Impact	A070909	Oversensing		
									A090801	Signal Artifact/Noise		
									A1702	Device-Device Incompatibility		
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A070909	Oversensing		
									A090801	Signal Artifact/Noise		
									A1702	Device-Device Incompatibility		
							F26	No Health Consequences or Impact	A070909	Oversensing		
									A090801	Signal Artifact/Noise		
									A1702	Device-Device Incompatibility		
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A070909	Oversensing		
									A090801	Signal Artifact/Noise		
									A1702	Device-Device Incompatibility		
							F26	No Health Consequences or Impact	A070909	Oversensing		
									A090801	Signal Artifact/Noise		
									A1702	Device-Device Incompatibility		
16386	2024	All other reportable incidents	47205	Cardiac pulse generator programmer	CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A09	Output Problem		
									A0902	Display or Visual Feedback Problem		
									A090801	Signal Artifact/Noise		
									A13	Communication or Transmission Problem		
							F23	Unexpected Medical Intervention	A09	Output Problem		
									A0902	Display or Visual Feedback Problem		
									A090801	Signal Artifact/Noise		
									A13	Communication or Transmission Problem		
16387	2024	All other reportable incidents	47265	Dual-chamber implantable pacemaker, rate-responsive	CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A0203	Defective Device	05/12/2024	05/12/2024
									A070504	Premature Discharge of Battery	05/12/2024	05/12/2024
									A0908	Incorrect, Inadequate or Imprecise Result or Readings	05/12/2024	05/12/2024
							F08	Hospitalization or Prolonged Hospitalization	A0203	Defective Device	05/12/2024	05/12/2024
									A070504	Premature Discharge of Battery	05/12/2024	05/12/2024
									A0908	Incorrect, Inadequate or Imprecise Result or Readings	05/12/2024	05/12/2024
							F1905	Device Revision or Replacement	A0203	Defective Device	05/12/2024	05/12/2024
									A070504	Premature Discharge of Battery	05/12/2024	05/12/2024
									A0908	Incorrect, Inadequate or Imprecise Result or Readings	05/12/2024	05/12/2024
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A0203	Defective Device	05/12/2024	05/12/2024
									A070504	Premature Discharge of Battery	05/12/2024	05/12/2024
									A0908	Incorrect, Inadequate or Imprecise Result or Readings	05/12/2024	05/12/2024
							F08	Hospitalization or Prolonged Hospitalization	A0203	Defective Device	05/12/2024	05/12/2024
									A070504	Premature Discharge of Battery	05/12/2024	05/12/2024
									A0908	Incorrect, Inadequate or Imprecise Result or Readings	05/12/2024	05/12/2024
							F1905	Device Revision or Replacement	A0203	Defective Device	05/12/2024	05/12/2024
									A070504	Premature Discharge of Battery	05/12/2024	05/12/2024
									A0908	Incorrect, Inadequate or Imprecise Result or Readings	05/12/2024	05/12/2024
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A0203	Defective Device	05/12/2024	05/12/2024
									A070504	Premature Discharge of Battery	05/12/2024	05/12/2024
									A0908	Incorrect, Inadequate or Imprecise Result or Readings	05/12/2024	05/12/2024
							F08	Hospitalization or Prolonged Hospitalization	A0203	Defective Device	05/12/2024	05/12/2024
									A070504	Premature Discharge of Battery	05/12/2024	05/12/2024
									A0908	Incorrect, Inadequate or Imprecise Result or Readings	05/12/2024	05/12/2024
							F1905	Device Revision or Replacement	A0203	Defective Device	05/12/2024	05/12/2024
									A070504	Premature Discharge of Battery	05/12/2024	05/12/2024
									A0908	Incorrect, Inadequate or Imprecise Result or Readings	05/12/2024	05/12/2024
16388	2024	All other reportable incidents	47265	Dual-chamber implantable pacemaker, rate-	CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A0712	Pacing Problem	10/07/2024	10/07/2024

				responsive					A0908	Incorrect, Inadequate or Imprecise Result or Readings	10/07/2024	10/07/2024
							F08	Hospitalization or Prolonged Hospitalization	A0712	Pacing Problem	10/07/2024	10/07/2024
									A0908	Incorrect, Inadequate or Imprecise Result or Readings	10/07/2024	10/07/2024
							F1001	Absence of Treatment	A0712	Pacing Problem	10/07/2024	10/07/2024
									A0908	Incorrect, Inadequate or Imprecise Result or Readings	10/07/2024	10/07/2024
							F1905	Device Revision or Replacement	A0712	Pacing Problem	10/07/2024	10/07/2024
									A0908	Incorrect, Inadequate or Imprecise Result or Readings	10/07/2024	10/07/2024
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A0712	Pacing Problem	10/07/2024	10/07/2024
									A0908	Incorrect, Inadequate or Imprecise Result or Readings	10/07/2024	10/07/2024
							F08	Hospitalization or Prolonged Hospitalization	A0712	Pacing Problem	10/07/2024	10/07/2024
									A0908	Incorrect, Inadequate or Imprecise Result or Readings	10/07/2024	10/07/2024
							F1001	Absence of Treatment	A0712	Pacing Problem	10/07/2024	10/07/2024
									A0908	Incorrect, Inadequate or Imprecise Result or Readings	10/07/2024	10/07/2024
							F1905	Device Revision or Replacement	A0712	Pacing Problem	10/07/2024	10/07/2024
									A0908	Incorrect, Inadequate or Imprecise Result or Readings	10/07/2024	10/07/2024
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A0712	Pacing Problem	10/07/2024	10/07/2024
									A0908	Incorrect, Inadequate or Imprecise Result or Readings	10/07/2024	10/07/2024
							F08	Hospitalization or Prolonged Hospitalization	A0712	Pacing Problem	10/07/2024	10/07/2024
									A0908	Incorrect, Inadequate or Imprecise Result or Readings	10/07/2024	10/07/2024
							F1001	Absence of Treatment	A0712	Pacing Problem	10/07/2024	10/07/2024
									A0908	Incorrect, Inadequate or Imprecise Result or Readings	10/07/2024	10/07/2024
							F1905	Device Revision or Replacement	A0712	Pacing Problem	10/07/2024	10/07/2024
									A0908	Incorrect, Inadequate or Imprecise Result or Readings	10/07/2024	10/07/2024
16389	2024	All other reportable incidents	47270	Cardiac resynchronization therapy implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
									A0908	Incorrect, Inadequate or Imprecise Result or Readings		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
									A0908	Incorrect, Inadequate or Imprecise Result or Readings		
16390	2024	All other reportable incidents	47265	Dual-chamber implantable pacemaker, rate-responsive	CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	24/12/2024	24/12/2024
									A0711	Interrogation Problem	24/12/2024	24/12/2024
							F1901	Additional Surgery	A070504	Premature Discharge of Battery	24/12/2024	24/12/2024
									A0711	Interrogation Problem	24/12/2024	24/12/2024
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	24/12/2024	24/12/2024
									A0711	Interrogation Problem	24/12/2024	24/12/2024
							F1901	Additional Surgery	A070504	Premature Discharge of Battery	24/12/2024	24/12/2024
									A0711	Interrogation Problem	24/12/2024	24/12/2024
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	24/12/2024	24/12/2024
									A0711	Interrogation Problem	24/12/2024	24/12/2024
							F1901	Additional Surgery	A070504	Premature Discharge of Battery	24/12/2024	24/12/2024
									A0711	Interrogation Problem	24/12/2024	24/12/2024
16391	2024	All other reportable incidents	17882	Physiologic-monitoring defibrillation system	CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A0908	Incorrect, Inadequate or Imprecise Result or Readings		
							F27	Problem Identified During Non-Clinical Procedure	A0908	Incorrect, Inadequate or Imprecise Result or Readings		
16392	2024	All other reportable incidents	35852	Single-chamber implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A0908	Incorrect, Inadequate or Imprecise Result or Readings		
							F26	No Health Consequences or Impact	A0908	Incorrect, Inadequate or Imprecise Result or Readings		
16393	2024	All other reportable incidents	37265	Dual-chamber implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A05	Mechanical Problem	16/12/2024	16/12/2024
							F26	No Health Consequences or Impact	A05	Mechanical Problem	16/12/2024	16/12/2024
16394	2024	All other reportable incidents	47265	Dual-chamber implantable pacemaker, rate-responsive	CT1685	Cardiac pacemakers	E011903	Syncope/Fainting	A070909	Oversensing	18/12/2024	18/12/2024
									A0712	Pacing Problem	18/12/2024	18/12/2024
									A0908	Incorrect, Inadequate or Imprecise Result or Readings	18/12/2024	18/12/2024
							F08	Hospitalization or Prolonged Hospitalization	A070909	Oversensing	18/12/2024	18/12/2024
									A0712	Pacing Problem	18/12/2024	18/12/2024
									A0908	Incorrect, Inadequate or Imprecise Result or Readings	18/12/2024	18/12/2024
							F1001	Absence of Treatment	A070909	Oversensing	18/12/2024	18/12/2024
									A0712	Pacing Problem	18/12/2024	18/12/2024
									A0908	Incorrect, Inadequate or Imprecise Result or Readings	18/12/2024	18/12/2024
							F1905	Device Revision or Replacement	A070909	Oversensing	18/12/2024	18/12/2024
									A0712	Pacing Problem	18/12/2024	18/12/2024
									A0908	Incorrect, Inadequate or Imprecise Result or Readings	18/12/2024	18/12/2024
					CT257	Implantable cardiac pacemakers	E011903	Syncope/Fainting	A070909	Oversensing	18/12/2024	18/12/2024
									A0712	Pacing Problem	18/12/2024	18/12/2024
									A0908	Incorrect, Inadequate or Imprecise Result or Readings	18/12/2024	18/12/2024
							F08	Hospitalization or Prolonged Hospitalization	A070909	Oversensing	18/12/2024	18/12/2024
									A0712	Pacing Problem	18/12/2024	18/12/2024
									A0908	Incorrect, Inadequate or Imprecise Result or Readings	18/12/2024	18/12/2024
							F1001	Absence of Treatment	A070909	Oversensing	18/12/2024	18/12/2024
									A0712	Pacing Problem	18/12/2024	18/12/2024
									A0908	Incorrect, Inadequate or Imprecise Result or Readings	18/12/2024	18/12/2024
							F1905	Device Revision or Replacement	A070909	Oversensing	18/12/2024	18/12/2024
									A0712	Pacing Problem	18/12/2024	18/12/2024

								A0908	Incorrect, Inadequate or Imprecise Result or Readings	18/12/2024	18/12/2024
					CT2866	Implantable cardiac electrical stimulators	E011903	Syncope/Fainting	A070909	Oversensing	18/12/2024	18/12/2024
									A0712	Pacing Problem	18/12/2024	18/12/2024
									A0908	Incorrect, Inadequate or Imprecise Result or Readings	18/12/2024	18/12/2024
							F08	Hospitalization or Prolonged Hospitalization	A070909	Oversensing	18/12/2024	18/12/2024
									A0712	Pacing Problem	18/12/2024	18/12/2024
									A0908	Incorrect, Inadequate or Imprecise Result or Readings	18/12/2024	18/12/2024
							F1001	Absence of Treatment	A070909	Oversensing	18/12/2024	18/12/2024
									A0712	Pacing Problem	18/12/2024	18/12/2024
									A0908	Incorrect, Inadequate or Imprecise Result or Readings	18/12/2024	18/12/2024
							F1905	Device Revision or Replacement	A070909	Oversensing	18/12/2024	18/12/2024
									A0712	Pacing Problem	18/12/2024	18/12/2024
									A0908	Incorrect, Inadequate or Imprecise Result or Readings	18/12/2024	18/12/2024
16395	2024	All other reportable incidents	37265	Dual-chamber implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E210401	Shock from Patient Lead(s)	A0709	Device Sensing Problem		
									A071302	Inappropriate/Inadequate Shock/Stimulation		
							F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A0709	Device Sensing Problem		
									A071302	Inappropriate/Inadequate Shock/Stimulation		
							F1203	Life Threatening Illness or Injury	A0709	Device Sensing Problem		
									A071302	Inappropriate/Inadequate Shock/Stimulation		
							F2307	Reprogramming of Device	A0709	Device Sensing Problem		
									A071302	Inappropriate/Inadequate Shock/Stimulation		
16396	2024	Death	37265	Dual-chamber implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E060110	Ventricular Fibrillation	A070910	Undersensing	08/12/2024	08/12/2024
							E0602	Cardiac Arrest	A070910	Undersensing	08/12/2024	08/12/2024
							E0717	Dyspnea	A070910	Undersensing	08/12/2024	08/12/2024
							E230101	Fever	A070910	Undersensing	08/12/2024	08/12/2024
							F02	Death	A070910	Undersensing	08/12/2024	08/12/2024
16397	2024	All other reportable incidents	60789	Intracardiac pacemaker	CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A051201	Device Dislodged or Dislocated		
							F26	No Health Consequences or Impact	A051201	Device Dislodged or Dislocated		
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A051201	Device Dislodged or Dislocated		
							F26	No Health Consequences or Impact	A051201	Device Dislodged or Dislocated		
					CT2818	Catheterization/Cannulation/Fluid path devices	E2403	No Clinical Signs, Symptoms or Conditions	A051201	Device Dislodged or Dislocated		
							F26	No Health Consequences or Impact	A051201	Device Dislodged or Dislocated		
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A051201	Device Dislodged or Dislocated		
							F26	No Health Consequences or Impact	A051201	Device Dislodged or Dislocated		
16398	2024	All other reportable incidents	47265	Dual-chamber implantable pacemaker, rate-responsive	CT1685	Cardiac pacemakers	E0112	Dizziness	A070504	Premature Discharge of Battery	10/12/2024	10/12/2024
									A0908	Incorrect, Inadequate or Imprecise Result or Readings	10/12/2024	10/12/2024
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	10/12/2024	10/12/2024
									A0908	Incorrect, Inadequate or Imprecise Result or Readings	10/12/2024	10/12/2024
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	10/12/2024	10/12/2024
									A0908	Incorrect, Inadequate or Imprecise Result or Readings	10/12/2024	10/12/2024
					CT257	Implantable cardiac pacemakers	E0112	Dizziness	A070504	Premature Discharge of Battery	10/12/2024	10/12/2024
									A0908	Incorrect, Inadequate or Imprecise Result or Readings	10/12/2024	10/12/2024
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	10/12/2024	10/12/2024
									A0908	Incorrect, Inadequate or Imprecise Result or Readings	10/12/2024	10/12/2024
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	10/12/2024	10/12/2024
									A0908	Incorrect, Inadequate or Imprecise Result or Readings	10/12/2024	10/12/2024
					CT2866	Implantable cardiac electrical stimulators	E0112	Dizziness	A070504	Premature Discharge of Battery	10/12/2024	10/12/2024
									A0908	Incorrect, Inadequate or Imprecise Result or Readings	10/12/2024	10/12/2024
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	10/12/2024	10/12/2024
									A0908	Incorrect, Inadequate or Imprecise Result or Readings	10/12/2024	10/12/2024
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	10/12/2024	10/12/2024
									A0908	Incorrect, Inadequate or Imprecise Result or Readings	10/12/2024	10/12/2024
16399	2024	All other reportable incidents	47265	Dual-chamber implantable pacemaker, rate-responsive	CT1685	Cardiac pacemakers	E060101	Asystole	A070909	Oversensing	17/10/2024	17/10/2024
									A0712	Pacing Problem	17/10/2024	17/10/2024
									A0908	Incorrect, Inadequate or Imprecise Result or Readings	17/10/2024	17/10/2024
							F08	Hospitalization or Prolonged Hospitalization	A070909	Oversensing	17/10/2024	17/10/2024
									A0712	Pacing Problem	17/10/2024	17/10/2024
									A0908	Incorrect, Inadequate or Imprecise Result or Readings	17/10/2024	17/10/2024
							F1001	Absence of Treatment	A070909	Oversensing	17/10/2024	17/10/2024
									A0712	Pacing Problem	17/10/2024	17/10/2024
									A0908	Incorrect, Inadequate or Imprecise Result or Readings	17/10/2024	17/10/2024
							F1905	Device Revision or Replacement	A070909	Oversensing	17/10/2024	17/10/2024
									A0712	Pacing Problem	17/10/2024	17/10/2024
									A0908	Incorrect, Inadequate or Imprecise Result or Readings	17/10/2024	17/10/2024
					CT257	Implantable cardiac pacemakers	E060101	Asystole	A070909	Oversensing	17/10/2024	17/10/2024
									A0712	Pacing Problem	17/10/2024	17/10/2024
									A0908	Incorrect, Inadequate or Imprecise Result or Readings	17/10/2024	17/10/2024
							F08	Hospitalization or Prolonged Hospitalization	A070909	Oversensing	17/10/2024	17/10/2024
									A0712	Pacing Problem	17/10/2024	17/10/2024
									A0908	Incorrect, Inadequate or Imprecise Result or Readings	17/10/2024	17/10/2024
							F1001	Absence of Treatment	A070909	Oversensing	17/10/2024	17/10/2024
									A0712	Pacing Problem	17/10/2024	17/10/2024

									A0908	Incorrect, Inadequate or Imprecise Result or Readings	17/10/2024	17/10/2024
							F1905	Device Revision or Replacement	A070909	Oversensing	17/10/2024	17/10/2024
									A0712	Pacing Problem	17/10/2024	17/10/2024
									A0908	Incorrect, Inadequate or Imprecise Result or Readings	17/10/2024	17/10/2024
					CT2866	Implantable cardiac electrical stimulators	E060101	Asystole	A070909	Oversensing	17/10/2024	17/10/2024
									A0712	Pacing Problem	17/10/2024	17/10/2024
									A0908	Incorrect, Inadequate or Imprecise Result or Readings	17/10/2024	17/10/2024
							F08	Hospitalization or Prolonged Hospitalization	A070909	Oversensing	17/10/2024	17/10/2024
									A0712	Pacing Problem	17/10/2024	17/10/2024
									A0908	Incorrect, Inadequate or Imprecise Result or Readings	17/10/2024	17/10/2024
							F1001	Absence of Treatment	A070909	Oversensing	17/10/2024	17/10/2024
									A0712	Pacing Problem	17/10/2024	17/10/2024
									A0908	Incorrect, Inadequate or Imprecise Result or Readings	17/10/2024	17/10/2024
							F1905	Device Revision or Replacement	A070909	Oversensing	17/10/2024	17/10/2024
									A0712	Pacing Problem	17/10/2024	17/10/2024
									A0908	Incorrect, Inadequate or Imprecise Result or Readings	17/10/2024	17/10/2024
16400	2024	All other reportable incidents	47265	Dual-chamber implantable pacemaker, rate-responsive	CT1685	Cardiac pacemakers	E2103	Device Overstimulation of Tissue	A070101	Failure to Capture	07/07/2025	07/07/2025
							F08	Hospitalization or Prolonged Hospitalization	A070101	Failure to Capture	07/07/2025	07/07/2025
							F1905	Device Revision or Replacement	A070101	Failure to Capture	07/07/2025	07/07/2025
							F2203	Imaging Required	A070101	Failure to Capture	07/07/2025	07/07/2025
					CT257	Implantable cardiac pacemakers	E2103	Device Overstimulation of Tissue	A070101	Failure to Capture	07/07/2025	07/07/2025
							F08	Hospitalization or Prolonged Hospitalization	A070101	Failure to Capture	07/07/2025	07/07/2025
							F1905	Device Revision or Replacement	A070101	Failure to Capture	07/07/2025	07/07/2025
							F2203	Imaging Required	A070101	Failure to Capture	07/07/2025	07/07/2025
					CT2866	Implantable cardiac electrical stimulators	E2103	Device Overstimulation of Tissue	A070101	Failure to Capture	07/07/2025	07/07/2025
							F08	Hospitalization or Prolonged Hospitalization	A070101	Failure to Capture	07/07/2025	07/07/2025
							F1905	Device Revision or Replacement	A070101	Failure to Capture	07/07/2025	07/07/2025
							F2203	Imaging Required	A070101	Failure to Capture	07/07/2025	07/07/2025
16401	2024	All other reportable incidents	47265	Dual-chamber implantable pacemaker, rate-responsive	CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
16402	2024	All other reportable incidents	37265	Dual-chamber implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem		
							F08	Hospitalization or Prolonged Hospitalization	A0705	Battery Problem		
							F12	Serious Injury/ Illness/ Impairment	A0705	Battery Problem		
							F1903	Device Explantation	A0705	Battery Problem		
16403	2024	All other reportable incidents	35852	Single-chamber implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	24/01/2025	24/01/2025
									A0908	Incorrect, Inadequate or Imprecise Result or Readings	24/01/2025	24/01/2025
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	24/01/2025	24/01/2025
									A0908	Incorrect, Inadequate or Imprecise Result or Readings	24/01/2025	24/01/2025
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	24/01/2025	24/01/2025
									A0908	Incorrect, Inadequate or Imprecise Result or Readings	24/01/2025	24/01/2025
16404	2024		47265	Dual-chamber implantable pacemaker, rate-responsive	CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A0712	Pacing Problem		
									A071206	Pacing Asynchronously		
							F1905	Device Revision or Replacement	A0712	Pacing Problem		
									A071206	Pacing Asynchronously		
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A0712	Pacing Problem		
									A071206	Pacing Asynchronously		
							F1905	Device Revision or Replacement	A0712	Pacing Problem		
									A071206	Pacing Asynchronously		
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A0712	Pacing Problem		
									A071206	Pacing Asynchronously		
							F1905	Device Revision or Replacement	A0712	Pacing Problem		
									A071206	Pacing Asynchronously		
16405	2024	All other reportable incidents	47265	Dual-chamber implantable pacemaker, rate-responsive	CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A071102	Failure to Interrogate	31/01/2025	31/01/2025
							F1901	Additional Surgery	A071102	Failure to Interrogate	31/01/2025	31/01/2025
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A071102	Failure to Interrogate	31/01/2025	31/01/2025
							F1901	Additional Surgery	A071102	Failure to Interrogate	31/01/2025	31/01/2025
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A071102	Failure to Interrogate	31/01/2025	31/01/2025
							F1901	Additional Surgery	A071102	Failure to Interrogate	31/01/2025	31/01/2025
16406	2024	All other reportable incidents	37265	Dual-chamber implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E0601	Arrhythmia	A070910	Undersensing		
									A071201	Failure to Convert Rhythm		
							F12	Serious Injury/ Illness/ Impairment	A070910	Undersensing		
									A071201	Failure to Convert Rhythm		
16407	2024	All other reportable incidents	35852	Single-chamber implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E210401	Shock from Patient Lead(s)	A070909	Oversensing		
									A071302	Inappropriate/Inadequate Shock/Stimulation		
							F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A070909	Oversensing		
									A071302	Inappropriate/Inadequate Shock/Stimulation		
16408	2024	All other reportable incidents	47270	Cardiac resynchronization therapy implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A070909	Oversensing		
									A0712	Pacing Problem		
									A072001	Electromagnetic Interference		
							F12	Serious Injury/ Illness/ Impairment	A070909	Oversensing		
									A0712	Pacing Problem		

									A072001	Electromagnetic Interference		
						F2307	Reprogramming of Device	A070909	Oversensing			
								A0712	Pacing Problem			
								A072001	Electromagnetic Interference			
						F26	No Health Consequences or Impact	A070909	Oversensing			
								A0712	Pacing Problem			
								A072001	Electromagnetic Interference			
16409	2024	All other reportable incidents	47270	Cardiac resynchronization therapy implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	20/01/2025	20/01/2025
									A0908	Incorrect, Inadequate or Imprecise Result or Readings	20/01/2025	20/01/2025
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	20/01/2025	20/01/2025
									A0908	Incorrect, Inadequate or Imprecise Result or Readings	20/01/2025	20/01/2025
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	20/01/2025	20/01/2025
									A0908	Incorrect, Inadequate or Imprecise Result or Readings	20/01/2025	20/01/2025
16410	2024	All other reportable incidents	35852	Single-chamber implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A090812	Missing Test Results	20/12/2024	20/12/2024
							F27	Problem Identified During Non-Clinical Procedure	A090812	Missing Test Results	20/12/2024	20/12/2024
16411	2024	All other reportable incidents	47265	Dual-chamber implantable pacemaker, rate-responsive	CT1685	Cardiac pacemakers	E0112	Dizziness	A070504	Premature Discharge of Battery	18/12/2024	18/12/2024
									A071102	Failure to Interrogate	18/12/2024	18/12/2024
									A071204	No Pacing	18/12/2024	18/12/2024
							E0601	Arrhythmia	A070504	Premature Discharge of Battery	18/12/2024	18/12/2024
									A071102	Failure to Interrogate	18/12/2024	18/12/2024
									A071204	No Pacing	18/12/2024	18/12/2024
							F1901	Additional Surgery	A070504	Premature Discharge of Battery	18/12/2024	18/12/2024
									A071102	Failure to Interrogate	18/12/2024	18/12/2024
									A071204	No Pacing	18/12/2024	18/12/2024
					CT257	Implantable cardiac pacemakers	E0112	Dizziness	A070504	Premature Discharge of Battery	18/12/2024	18/12/2024
									A071102	Failure to Interrogate	18/12/2024	18/12/2024
									A071204	No Pacing	18/12/2024	18/12/2024
							E0601	Arrhythmia	A070504	Premature Discharge of Battery	18/12/2024	18/12/2024
									A071102	Failure to Interrogate	18/12/2024	18/12/2024
									A071204	No Pacing	18/12/2024	18/12/2024
							F1901	Additional Surgery	A070504	Premature Discharge of Battery	18/12/2024	18/12/2024
									A071102	Failure to Interrogate	18/12/2024	18/12/2024
									A071204	No Pacing	18/12/2024	18/12/2024
					CT2866	Implantable cardiac electrical stimulators	E0112	Dizziness	A070504	Premature Discharge of Battery	18/12/2024	18/12/2024
									A071102	Failure to Interrogate	18/12/2024	18/12/2024
									A071204	No Pacing	18/12/2024	18/12/2024
							E0601	Arrhythmia	A070504	Premature Discharge of Battery	18/12/2024	18/12/2024
									A071102	Failure to Interrogate	18/12/2024	18/12/2024
									A071204	No Pacing	18/12/2024	18/12/2024
							F1901	Additional Surgery	A070504	Premature Discharge of Battery	18/12/2024	18/12/2024
									A071102	Failure to Interrogate	18/12/2024	18/12/2024
									A071204	No Pacing	18/12/2024	18/12/2024
16412	2024	All other reportable incidents	17882	Physiologic-monitoring defibrillation system	CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A0801	Failure to Calibrate		
							F27	Problem Identified During Non-Clinical Procedure	A0801	Failure to Calibrate		
16413	2024	All other reportable incidents	17882	Physiologic-monitoring defibrillation system	CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A0908	Incorrect, Inadequate or Imprecise Result or Readings		
							F26	No Health Consequences or Impact	A0908	Incorrect, Inadequate or Imprecise Result or Readings		
16414	2024	All other reportable incidents	47265	Dual-chamber implantable pacemaker, rate-responsive	CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A072202	Low Impedance		
									A160105	False Alarm		
							F26	No Health Consequences or Impact	A072202	Low Impedance		
									A160105	False Alarm		
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A072202	Low Impedance		
									A160105	False Alarm		
							F26	No Health Consequences or Impact	A072202	Low Impedance		
									A160105	False Alarm		
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A072202	Low Impedance		
									A160105	False Alarm		
							F26	No Health Consequences or Impact	A072202	Low Impedance		
									A160105	False Alarm		
16415	2024	All other reportable incidents	47265	Dual-chamber implantable pacemaker, rate-responsive	CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A0908	Incorrect, Inadequate or Imprecise Result or Readings	29/11/2024	29/11/2024
							F08	Hospitalization or Prolonged Hospitalization	A0908	Incorrect, Inadequate or Imprecise Result or Readings	29/11/2024	29/11/2024
							F1905	Device Revision or Replacement	A0908	Incorrect, Inadequate or Imprecise Result or Readings	29/11/2024	29/11/2024
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A0908	Incorrect, Inadequate or Imprecise Result or Readings	29/11/2024	29/11/2024
							F08	Hospitalization or Prolonged Hospitalization	A0908	Incorrect, Inadequate or Imprecise Result or Readings	29/11/2024	29/11/2024
							F1905	Device Revision or Replacement	A0908	Incorrect, Inadequate or Imprecise Result or Readings	29/11/2024	29/11/2024
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A0908	Incorrect, Inadequate or Imprecise Result or Readings	29/11/2024	29/11/2024
							F08	Hospitalization or Prolonged Hospitalization	A0908	Incorrect, Inadequate or Imprecise Result or Readings	29/11/2024	29/11/2024
							F1905	Device Revision or Replacement	A0908	Incorrect, Inadequate or Imprecise Result or Readings	29/11/2024	29/11/2024
16416	2024	All other reportable incidents	47265	Dual-chamber implantable pacemaker, rate-responsive	CT1685	Cardiac pacemakers	E011902	Presyncope	A1304	Telemetry Discrepancy		
							F08	Hospitalization or Prolonged Hospitalization	A1304	Telemetry Discrepancy		
					CT257	Implantable cardiac pacemakers	E011902	Presyncope	A1304	Telemetry Discrepancy		
							F08	Hospitalization or Prolonged Hospitalization	A1304	Telemetry Discrepancy		
					CT2866	Implantable cardiac electrical stimulators	E011902	Presyncope	A1304	Telemetry Discrepancy		
							F08	Hospitalization or Prolonged Hospitalization	A1304	Telemetry Discrepancy		
16417	2024	All other reportable incidents	47265	Dual-chamber implantable pacemaker, rate-responsive	CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A0908	Incorrect, Inadequate or Imprecise Result or Readings	29/11/2024	29/11/2024
							F08	Hospitalization or Prolonged Hospitalization	A0908	Incorrect, Inadequate or Imprecise Result or Readings	29/11/2024	29/11/2024

							F1905	Device Revision or Replacement	A0908	Incorrect, Inadequate or Imprecise Result or Readings	29/11/2024	29/11/2024
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A0908	Incorrect, Inadequate or Imprecise Result or Readings	29/11/2024	29/11/2024
							F08	Hospitalization or Prolonged Hospitalization	A0908	Incorrect, Inadequate or Imprecise Result or Readings	29/11/2024	29/11/2024
							F1905	Device Revision or Replacement	A0908	Incorrect, Inadequate or Imprecise Result or Readings	29/11/2024	29/11/2024
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A0908	Incorrect, Inadequate or Imprecise Result or Readings	29/11/2024	29/11/2024
							F08	Hospitalization or Prolonged Hospitalization	A0908	Incorrect, Inadequate or Imprecise Result or Readings	29/11/2024	29/11/2024
							F1905	Device Revision or Replacement	A0908	Incorrect, Inadequate or Imprecise Result or Readings	29/11/2024	29/11/2024
16418	2024	All other reportable incidents	17882	Physiologic-monitoring defibrillation system	CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A1603	Failure of Device to Self-Test		
							F26	No Health Consequences or Impact	A1603	Failure of Device to Self-Test		
16419	2024		60789	Intracardiac pacemaker	CT1685	Cardiac pacemakers	E0601	Arrhythmia	A24	No Apparent Device Problem		
							F02	Death	A24	No Apparent Device Problem		
					CT257	Implantable cardiac pacemakers	E0601	Arrhythmia	A24	No Apparent Device Problem		
							F02	Death	A24	No Apparent Device Problem		
					CT2818	Catheterization/Cannulation/Fluid path devices	E0601	Arrhythmia	A24	No Apparent Device Problem		
							F02	Death	A24	No Apparent Device Problem		
					CT2866	Implantable cardiac electrical stimulators	E0601	Arrhythmia	A24	No Apparent Device Problem		
							F02	Death	A24	No Apparent Device Problem		
16420	2024	All other reportable incidents	17882	Physiologic-monitoring defibrillation system	CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A0719	Unexpected Shutdown		
							F27	Problem Identified During Non-Clinical Procedure	A0719	Unexpected Shutdown		
16421	2024	All other reportable incidents	47205	Cardiac pulse generator programmer	CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A1104	Computer Operating System Problem		
							F23	Unexpected Medical Intervention	A1104	Computer Operating System Problem		
16422	2024	All other reportable incidents	17882	Physiologic-monitoring defibrillation system	CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A071301	Failure to Deliver Shock/Stimulation		
							F29	Death not Related to Reported Adverse Event	A071301	Failure to Deliver Shock/Stimulation		
16423	2024	All other reportable incidents	37265	Dual-chamber implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A072201	High Impedance		
									A090801	Signal Artifact/Noise		
									A1301	Failure to Read Input Signal		
							F26	No Health Consequences or Impact	A072201	High Impedance		
									A090801	Signal Artifact/Noise		
									A1301	Failure to Read Input Signal		
16424	2024	All other reportable incidents	47265	Dual-chamber implantable pacemaker, rate-responsive	CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A0711	Interrogation Problem	16/12/2024	16/12/2024
							F08	Hospitalization or Prolonged Hospitalization	A0711	Interrogation Problem	16/12/2024	16/12/2024
							F1905	Device Revision or Replacement	A0711	Interrogation Problem	16/12/2024	16/12/2024
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A0711	Interrogation Problem	16/12/2024	16/12/2024
							F08	Hospitalization or Prolonged Hospitalization	A0711	Interrogation Problem	16/12/2024	16/12/2024
							F1905	Device Revision or Replacement	A0711	Interrogation Problem	16/12/2024	16/12/2024
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A0711	Interrogation Problem	16/12/2024	16/12/2024
							F08	Hospitalization or Prolonged Hospitalization	A0711	Interrogation Problem	16/12/2024	16/12/2024
							F1905	Device Revision or Replacement	A0711	Interrogation Problem	16/12/2024	16/12/2024
16425	2024	All other reportable incidents	35852	Single-chamber implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	08/01/2025	08/01/2025
							F1901	Additional Surgery	A070504	Premature Discharge of Battery	08/01/2025	08/01/2025
16426	2024		17882	Physiologic-monitoring defibrillation system	CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem		
							F27	Problem Identified During Non-Clinical Procedure	A24	No Apparent Device Problem		
16427	2024	All other reportable incidents	47265	Dual-chamber implantable pacemaker, rate-responsive	CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
16428	2024	All other reportable incidents	47265	Dual-chamber implantable pacemaker, rate-responsive	CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A0908	Incorrect, Inadequate or Imprecise Result or Readings		
							F26	No Health Consequences or Impact	A0908	Incorrect, Inadequate or Imprecise Result or Readings		
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A0908	Incorrect, Inadequate or Imprecise Result or Readings		
							F26	No Health Consequences or Impact	A0908	Incorrect, Inadequate or Imprecise Result or Readings		
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A0908	Incorrect, Inadequate or Imprecise Result or Readings		
							F26	No Health Consequences or Impact	A0908	Incorrect, Inadequate or Imprecise Result or Readings		
16429	2024	All other reportable incidents	47265	Dual-chamber implantable pacemaker, rate-responsive	CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A1605	Reset Problem		
							F26	No Health Consequences or Impact	A1605	Reset Problem		
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A1605	Reset Problem		
							F26	No Health Consequences or Impact	A1605	Reset Problem		
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A1605	Reset Problem		
							F26	No Health Consequences or Impact	A1605	Reset Problem		
16430	2024	All other reportable incidents	17882	Physiologic-monitoring defibrillation system	CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A0719	Unexpected Shutdown		
							F26	No Health Consequences or Impact	A0719	Unexpected Shutdown		
16431	2024	All other reportable incidents	36106	Implantable pacemaker analyser, intraoperative	CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A11	Computer Software Problem		
							F26	No Health Consequences or Impact	A11	Computer Software Problem		
16432	2024	All other reportable incidents	36106	Implantable pacemaker analyser, intraoperative	CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A11	Computer Software Problem		
							F26	No Health Consequences or Impact	A11	Computer Software Problem		
16433	2024	All other reportable incidents	17882	Physiologic-monitoring defibrillation system	CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A0719	Unexpected Shutdown		
							F27	Problem Identified During Non-Clinical Procedure	A0719	Unexpected Shutdown		
16434	2024	All other reportable incidents	47265	Dual-chamber implantable pacemaker, rate-responsive	CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		

							F26	No Health Consequences or Impact	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
16435	2025	All other reportable incidents	35852	Single-chamber implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A160602	Premature End-of-Life Indicator		
							F1905	Device Revision or Replacement	A160602	Premature End-of-Life Indicator		
16436	2025		45966	Transthoracic pacing lead	CT1684	Cardiac pacemakers and associated devices	F26	No Health Consequences or Impact	A0202	Defective Component		
16437	2025	All other reportable incidents	17882	Physiologic-monitoring defibrillation system	CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A150105	Key or Button Unresponsive/not Working		
							F27	Problem Identified During Non-Clinical Procedure	A150105	Key or Button Unresponsive/not Working		
16438	2025	All other reportable incidents			CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	20/12/2024	20/12/2024
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery	20/12/2024	20/12/2024
16439	2025	All other reportable incidents	47270	Cardiac resynchronization therapy implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E210401	Shock from Patient Lead(s)	A072201	High Impedance		
							F17	Sedation	A072201	High Impedance		
							F22	Unexpected Diagnostic Intervention	A072201	High Impedance		
16440	2025	All other reportable incidents	37265	Dual-chamber implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E060101	Asystole	A070909	Oversensing		
									A0712	Pacing Problem		
									A090801	Signal Artifact/Noise		
							F1001	Absence of Treatment	A070909	Oversensing		
									A0712	Pacing Problem		
									A090801	Signal Artifact/Noise		
16441	2025	All other reportable incidents	47270	Cardiac resynchronization therapy implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A070909	Oversensing		
									A071204	No Pacing		
							F22	Unexpected Diagnostic Intervention	A070909	Oversensing		
									A071204	No Pacing		
16442	2025	All other reportable incidents	35852	Single-chamber implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A160601	Premature Elective Replacement Indicator	27/12/2024	27/12/2024
							F1905	Device Revision or Replacement	A160601	Premature Elective Replacement Indicator	27/12/2024	27/12/2024
16443	2025	All other reportable incidents	36106	Implantable pacemaker analyser, intraoperative	CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A11	Computer Software Problem		
							F26	No Health Consequences or Impact	A11	Computer Software Problem		
16444	2025	All other reportable incidents	36106	Implantable pacemaker analyser, intraoperative	CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A090201	Device Displays Incorrect Message		
							F26	No Health Consequences or Impact	A090201	Device Displays Incorrect Message		
16445	2025	All other reportable incidents	37265	Dual-chamber implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E210401	Shock from Patient Lead(s)	A0709	Device Sensing Problem		
									A0712	Pacing Problem		
									A071302	Inappropriate/Inadequate Shock/Stimulation		
							F08	Hospitalization or Prolonged Hospitalization	A0709	Device Sensing Problem		
									A0712	Pacing Problem		
									A071302	Inappropriate/Inadequate Shock/Stimulation		
							F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A0709	Device Sensing Problem		
									A0712	Pacing Problem		
									A071302	Inappropriate/Inadequate Shock/Stimulation		
							F1203	Life Threatening Illness or Injury	A0709	Device Sensing Problem		
									A0712	Pacing Problem		
									A071302	Inappropriate/Inadequate Shock/Stimulation		
							F2307	Reprogramming of Device	A0709	Device Sensing Problem		
									A0712	Pacing Problem		
									A071302	Inappropriate/Inadequate Shock/Stimulation		
16446	2025	All other reportable incidents	37265	Dual-chamber implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E060109	Tachycardia	A0712	Pacing Problem		
							E060110	Ventricular Fibrillation	A0712	Pacing Problem		
							E1610	Head Injury	A0712	Pacing Problem		
							F08	Hospitalization or Prolonged Hospitalization	A0712	Pacing Problem		
							F1203	Life Threatening Illness or Injury	A0712	Pacing Problem		
							F2303	Medication Required	A0712	Pacing Problem		
							F2307	Reprogramming of Device	A0712	Pacing Problem		
16447	2025	All other reportable incidents	36106	Implantable pacemaker analyser, intraoperative	CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A11	Computer Software Problem		
							F26	No Health Consequences or Impact	A11	Computer Software Problem		
16448	2025	All other reportable incidents	36106	Implantable pacemaker analyser, intraoperative	CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A11	Computer Software Problem		
							F26	No Health Consequences or Impact	A11	Computer Software Problem		
16449	2025	All other reportable incidents	36106	Implantable pacemaker analyser, intraoperative	CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A11	Computer Software Problem		
							F26	No Health Consequences or Impact	A11	Computer Software Problem		
16450	2025	All other reportable incidents	36106	Implantable pacemaker analyser, intraoperative	CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A11	Computer Software Problem		
							F26	No Health Consequences or Impact	A11	Computer Software Problem		
16451	2025	All other reportable incidents	47267	Single-chamber implantable pacemaker, rate-responsive	CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
16452	2025	All other reportable incidents	47270	Cardiac resynchronization therapy implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem	05/02/2025	05/02/2025
									A070504	Premature Discharge of Battery	05/02/2025	05/02/2025
							F08	Hospitalization or Prolonged Hospitalization	A0705	Battery Problem	05/02/2025	05/02/2025
									A070504	Premature Discharge of Battery	05/02/2025	05/02/2025
							F1905	Device Revision or Replacement	A0705	Battery Problem	05/02/2025	05/02/2025
									A070504	Premature Discharge of Battery	05/02/2025	05/02/2025
16453	2025	All other reportable incidents	47270	Cardiac resynchronization therapy implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	23/12/2024	23/12/2024
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	23/12/2024	23/12/2024
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	23/12/2024	23/12/2024
16454	2025	All other reportable incidents	47267	Single-chamber implantable pacemaker, rate-responsive	CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A0908	Incorrect, Inadequate or Imprecise Result or Readings		
							F08	Hospitalization or Prolonged Hospitalization	A0908	Incorrect, Inadequate or Imprecise Result or Readings		
							F1905	Device Revision or Replacement	A0908	Incorrect, Inadequate or Imprecise Result or Readings		
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A0908	Incorrect, Inadequate or Imprecise Result or Readings		
							F08	Hospitalization or Prolonged Hospitalization	A0908	Incorrect, Inadequate or Imprecise Result or Readings		

							F1905	Device Revision or Replacement	A0908	Incorrect, Inadequate or Imprecise Result or Readings			
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A0908	Incorrect, Inadequate or Imprecise Result or Readings			
							F08	Hospitalization or Prolonged Hospitalization	A0908	Incorrect, Inadequate or Imprecise Result or Readings			
							F1905	Device Revision or Replacement	A0908	Incorrect, Inadequate or Imprecise Result or Readings			
16455	2025	All other reportable incidents			CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070909	Oversensing			
									A090801	Signal Artifact/Noise			
									A1301	Failure to Read Input Signal			
							F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A070909	Oversensing			
									A090801	Signal Artifact/Noise			
									A1301	Failure to Read Input Signal			
							F2203	Imaging Required	A070909	Oversensing			
									A090801	Signal Artifact/Noise			
									A1301	Failure to Read Input Signal			
16456	2025	All other reportable incidents	47267	Single-chamber implantable pacemaker, rate-responsive	CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A0908	Incorrect, Inadequate or Imprecise Result or Readings			
							F08	Hospitalization or Prolonged Hospitalization	A0908	Incorrect, Inadequate or Imprecise Result or Readings			
							F1905	Device Revision or Replacement	A0908	Incorrect, Inadequate or Imprecise Result or Readings			
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A0908	Incorrect, Inadequate or Imprecise Result or Readings			
							F08	Hospitalization or Prolonged Hospitalization	A0908	Incorrect, Inadequate or Imprecise Result or Readings			
							F1905	Device Revision or Replacement	A0908	Incorrect, Inadequate or Imprecise Result or Readings			
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A0908	Incorrect, Inadequate or Imprecise Result or Readings			
							F08	Hospitalization or Prolonged Hospitalization	A0908	Incorrect, Inadequate or Imprecise Result or Readings			
							F1905	Device Revision or Replacement	A0908	Incorrect, Inadequate or Imprecise Result or Readings			
16457	2025	All other reportable incidents	47267	Single-chamber implantable pacemaker, rate-responsive			E2403	No Clinical Signs, Symptoms or Conditions	A0908	Incorrect, Inadequate or Imprecise Result or Readings			
							F08	Hospitalization or Prolonged Hospitalization	A0908	Incorrect, Inadequate or Imprecise Result or Readings			
							F1905	Device Revision or Replacement	A0908	Incorrect, Inadequate or Imprecise Result or Readings			
16458	2025	All other reportable incidents			CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery			
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery			
16459	2025	All other reportable incidents	47270	Cardiac resynchronization therapy implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E210401	Shock from Patient Lead(s)	A070909	Oversensing	20/02/2025	20/02/2025	
									A071302	Inappropriate/Inadequate Shock/Stimulation	20/02/2025	20/02/2025	
							F08	Hospitalization or Prolonged Hospitalization	A070909	Oversensing	20/02/2025	20/02/2025	
									A071302	Inappropriate/Inadequate Shock/Stimulation	20/02/2025	20/02/2025	
							F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A070909	Oversensing	20/02/2025	20/02/2025	
									A071302	Inappropriate/Inadequate Shock/Stimulation	20/02/2025	20/02/2025	
							F1905	Device Revision or Replacement	A070909	Oversensing	20/02/2025	20/02/2025	
									A071302	Inappropriate/Inadequate Shock/Stimulation	20/02/2025	20/02/2025	
							F23	Unexpected Medical Intervention	A070909	Oversensing	20/02/2025	20/02/2025	
									A071302	Inappropriate/Inadequate Shock/Stimulation	20/02/2025	20/02/2025	
16460	2025	All other reportable incidents	47270	Cardiac resynchronization therapy implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E210401	Shock from Patient Lead(s)	A0712	Pacing Problem			
									A071302	Inappropriate/Inadequate Shock/Stimulation			
							F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A0712	Pacing Problem			
									A071302	Inappropriate/Inadequate Shock/Stimulation			
16461	2025	All other reportable incidents			CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070101	Failure to Capture			
									A070504	Premature Discharge of Battery			
									A072201	High Impedance			
							F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A070101	Failure to Capture			
									A070504	Premature Discharge of Battery			
									A072201	High Impedance			
16462	2025	All other reportable incidents	17882	Physiologic-monitoring defibrillation system	CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070803	Failure to Power Up			
							F27	Problem Identified During Non-Clinical Procedure	A070803	Failure to Power Up			
16463	2025	All other reportable incidents	36106	Implantable pacemaker analyser, intraoperative	CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A11	Computer Software Problem			
							F26	No Health Consequences or Impact	A11	Computer Software Problem			
16464	2025	All other reportable incidents	36106	Implantable pacemaker analyser, intraoperative	CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A11	Computer Software Problem			
							F26	No Health Consequences or Impact	A11	Computer Software Problem			
16465	2025	All other reportable incidents	36106	Implantable pacemaker analyser, intraoperative	CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A11	Computer Software Problem			
							F26	No Health Consequences or Impact	A11	Computer Software Problem			
16466	2025	All other reportable incidents	47265	Dual-chamber implantable pacemaker, rate-responsive	CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A0908	Incorrect, Inadequate or Imprecise Result or Readings			
							F26	No Health Consequences or Impact	A0908	Incorrect, Inadequate or Imprecise Result or Readings			
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A0908	Incorrect, Inadequate or Imprecise Result or Readings			
							F26	No Health Consequences or Impact	A0908	Incorrect, Inadequate or Imprecise Result or Readings			
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A0908	Incorrect, Inadequate or Imprecise Result or Readings			
							F26	No Health Consequences or Impact	A0908	Incorrect, Inadequate or Imprecise Result or Readings			
16467	2025	All other reportable incidents	47265	Dual-chamber implantable pacemaker, rate-responsive	CT1685	Cardiac pacemakers	E060101	Asystole	A070909	Oversensing	24/10/2024	24/10/2024	
									A0712	Pacing Problem	24/10/2024	24/10/2024	
									A090801	Signal Artifact/Noise	24/10/2024	24/10/2024	
							F1001	Absence of Treatment	A070909	Oversensing	24/10/2024	24/10/2024	
									A0712	Pacing Problem	24/10/2024	24/10/2024	
									A090801	Signal Artifact/Noise	24/10/2024	24/10/2024	
					CT257	Implantable cardiac pacemakers	E060101	Asystole	A070909	Oversensing	24/10/2024	24/10/2024	
									A0712	Pacing Problem	24/10/2024	24/10/2024	
									A090801	Signal Artifact/Noise	24/10/2024	24/10/2024	
							F1001	Absence of Treatment	A070909	Oversensing	24/10/2024	24/10/2024	

									A0712	Pacing Problem	24/10/2024	24/10/2024
									A090801	Signal Artifact/Noise	24/10/2024	24/10/2024
					CT2866	Implantable cardiac electrical stimulators	E060101	Asystole	A070909	Oversensing	24/10/2024	24/10/2024
									A0712	Pacing Problem	24/10/2024	24/10/2024
									A090801	Signal Artifact/Noise	24/10/2024	24/10/2024
							F1001	Absence of Treatment	A070909	Oversensing	24/10/2024	24/10/2024
									A0712	Pacing Problem	24/10/2024	24/10/2024
									A090801	Signal Artifact/Noise	24/10/2024	24/10/2024
16468	2025	All other reportable incidents	47270	Cardiac resynchronization therapy implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A070909	Oversensing		
									A0712	Pacing Problem		
							F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A070909	Oversensing		
									A0712	Pacing Problem		
16469	2025	All other reportable incidents	47263	Cardiac resynchronization therapy implantable pacemaker	CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A0908	Incorrect, Inadequate or Imprecise Result or Readings	06/08/2024	06/08/2024
							F08	Hospitalization or Prolonged Hospitalization	A0908	Incorrect, Inadequate or Imprecise Result or Readings	06/08/2024	06/08/2024
							F1905	Device Revision or Replacement	A0908	Incorrect, Inadequate or Imprecise Result or Readings	06/08/2024	06/08/2024
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A0908	Incorrect, Inadequate or Imprecise Result or Readings	06/08/2024	06/08/2024
							F08	Hospitalization or Prolonged Hospitalization	A0908	Incorrect, Inadequate or Imprecise Result or Readings	06/08/2024	06/08/2024
							F1905	Device Revision or Replacement	A0908	Incorrect, Inadequate or Imprecise Result or Readings	06/08/2024	06/08/2024
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A0908	Incorrect, Inadequate or Imprecise Result or Readings	06/08/2024	06/08/2024
							F08	Hospitalization or Prolonged Hospitalization	A0908	Incorrect, Inadequate or Imprecise Result or Readings	06/08/2024	06/08/2024
							F1905	Device Revision or Replacement	A0908	Incorrect, Inadequate or Imprecise Result or Readings	06/08/2024	06/08/2024
16470	2025	All other reportable incidents	37265	Dual-chamber implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A070102	High Capture Threshold		
									A0705	Battery Problem		
							F26	No Health Consequences or Impact	A070102	High Capture Threshold		
									A0705	Battery Problem		
16471	2025	All other reportable incidents	17882	Physiologic-monitoring defibrillation system	CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A0906	Gas Output Problem		
							F27	Problem Identified During Non-Clinical Procedure	A0906	Gas Output Problem		
16472	2025	All other reportable incidents	47267	Single-chamber implantable pacemaker, rate-responsive	CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
16473	2025	Unanticipated serious deterioration in state of health	35224	External pacemaker, epicardial pacing	CT1685	Cardiac pacemakers	E060101	Asystole	A0203	Defective Device		
									A070801	Complete Loss of Power		
									A0719	Unexpected Shutdown		
									A1601	Device Alarm System		
							F0801	Intensive Care	A0203	Defective Device		
									A070801	Complete Loss of Power		
									A0719	Unexpected Shutdown		
									A1601	Device Alarm System		
16474	2025	All other reportable incidents	36106	Implantable pacemaker analyser, intraoperative	CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A11	Computer Software Problem		
									A1107	Data Problem		
							F27	Problem Identified During Non-Clinical Procedure	A11	Computer Software Problem		
									A1107	Data Problem		
16475	2025	All other reportable incidents	36106	Implantable pacemaker analyser, intraoperative	CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A11	Computer Software Problem		
									A1107	Data Problem		
							F27	Problem Identified During Non-Clinical Procedure	A11	Computer Software Problem		
									A1107	Data Problem		
16476	2025	All other reportable incidents	47265	Dual-chamber implantable pacemaker, rate-responsive	CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A0908	Incorrect, Inadequate or Imprecise Result or Readings	16/01/2025	16/01/2025
							F08	Hospitalization or Prolonged Hospitalization	A0908	Incorrect, Inadequate or Imprecise Result or Readings	16/01/2025	16/01/2025
							F1905	Device Revision or Replacement	A0908	Incorrect, Inadequate or Imprecise Result or Readings	16/01/2025	16/01/2025
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A0908	Incorrect, Inadequate or Imprecise Result or Readings	16/01/2025	16/01/2025
							F08	Hospitalization or Prolonged Hospitalization	A0908	Incorrect, Inadequate or Imprecise Result or Readings	16/01/2025	16/01/2025
							F1905	Device Revision or Replacement	A0908	Incorrect, Inadequate or Imprecise Result or Readings	16/01/2025	16/01/2025
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A0908	Incorrect, Inadequate or Imprecise Result or Readings	16/01/2025	16/01/2025
							F08	Hospitalization or Prolonged Hospitalization	A0908	Incorrect, Inadequate or Imprecise Result or Readings	16/01/2025	16/01/2025
							F1905	Device Revision or Replacement	A0908	Incorrect, Inadequate or Imprecise Result or Readings	16/01/2025	16/01/2025
16477	2025	All other reportable incidents	35852	Single-chamber implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem		
							F24	Insufficient Health Impact Information	A0705	Battery Problem		
16478	2025	All other reportable incidents	35852	Single-chamber implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem	29/10/2024	29/10/2024
									A070504	Premature Discharge of Battery	29/10/2024	29/10/2024
									A1304	Telemetry Discrepancy	29/10/2024	29/10/2024
							F08	Hospitalization or Prolonged Hospitalization	A0705	Battery Problem	29/10/2024	29/10/2024
									A070504	Premature Discharge of Battery	29/10/2024	29/10/2024
									A1304	Telemetry Discrepancy	29/10/2024	29/10/2024
							F1905	Device Revision or Replacement	A0705	Battery Problem	29/10/2024	29/10/2024
									A070504	Premature Discharge of Battery	29/10/2024	29/10/2024
									A1304	Telemetry Discrepancy	29/10/2024	29/10/2024
16479	2025	All other reportable incidents	47265	Dual-chamber implantable pacemaker, rate-responsive	CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A0908	Incorrect, Inadequate or Imprecise Result or Readings	05/01/2025	05/01/2025
							F08	Hospitalization or Prolonged Hospitalization	A0908	Incorrect, Inadequate or Imprecise Result or Readings	05/01/2025	05/01/2025
							F1905	Device Revision or Replacement	A0908	Incorrect, Inadequate or Imprecise Result or Readings	05/01/2025	05/01/2025

					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A0908	Incorrect, Inadequate or Imprecise Result or Readings	05/01/2025	05/01/2025
							F08	Hospitalization or Prolonged Hospitalization	A0908	Incorrect, Inadequate or Imprecise Result or Readings	05/01/2025	05/01/2025
							F1905	Device Revision or Replacement	A0908	Incorrect, Inadequate or Imprecise Result or Readings	05/01/2025	05/01/2025
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A0908	Incorrect, Inadequate or Imprecise Result or Readings	05/01/2025	05/01/2025
							F08	Hospitalization or Prolonged Hospitalization	A0908	Incorrect, Inadequate or Imprecise Result or Readings	05/01/2025	05/01/2025
							F1905	Device Revision or Replacement	A0908	Incorrect, Inadequate or Imprecise Result or Readings	05/01/2025	05/01/2025
16480	2025	All other reportable incidents	47270	Cardiac resynchronization therapy implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem	06/05/2025	06/05/2025
									A070504	Premature Discharge of Battery	06/05/2025	06/05/2025
									A0908	Incorrect, Inadequate or Imprecise Result or Readings	06/05/2025	06/05/2025
							F08	Hospitalization or Prolonged Hospitalization	A0705	Battery Problem	06/05/2025	06/05/2025
									A070504	Premature Discharge of Battery	06/05/2025	06/05/2025
									A0908	Incorrect, Inadequate or Imprecise Result or Readings	06/05/2025	06/05/2025
							F1905	Device Revision or Replacement	A0705	Battery Problem	06/05/2025	06/05/2025
									A070504	Premature Discharge of Battery	06/05/2025	06/05/2025
									A0908	Incorrect, Inadequate or Imprecise Result or Readings	06/05/2025	06/05/2025
16481	2025	All other reportable incidents	17882	Physiologic-monitoring defibrillation system	CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A150105	Key or Button Unresponsive/not Working		
							F27	Problem Identified During Non-Clinical Procedure	A150105	Key or Button Unresponsive/not Working		
16482	2025	All other reportable incidents	47265	Dual-chamber implantable pacemaker, rate-responsive	CT1685	Cardiac pacemakers	E060101	Asystole	A070101	Failure to Capture		
									A070102	High Capture Threshold		
							F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A070101	Failure to Capture		
									A070102	High Capture Threshold		
					CT257	Implantable cardiac pacemakers	E060101	Asystole	A070101	Failure to Capture		
									A070102	High Capture Threshold		
							F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A070101	Failure to Capture		
									A070102	High Capture Threshold		
					CT2866	Implantable cardiac electrical stimulators	E060101	Asystole	A070101	Failure to Capture		
									A070102	High Capture Threshold		
							F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A070101	Failure to Capture		
									A070102	High Capture Threshold		
16483	2025	All other reportable incidents	47265	Dual-chamber implantable pacemaker, rate-responsive	CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A0709	Device Sensing Problem	14/08/2024	14/08/2024
									A0712	Pacing Problem	14/08/2024	14/08/2024
							F26	No Health Consequences or Impact	A0709	Device Sensing Problem	14/08/2024	14/08/2024
									A0712	Pacing Problem	14/08/2024	14/08/2024
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A0709	Device Sensing Problem	14/08/2024	14/08/2024
									A0712	Pacing Problem	14/08/2024	14/08/2024
							F26	No Health Consequences or Impact	A0709	Device Sensing Problem	14/08/2024	14/08/2024
									A0712	Pacing Problem	14/08/2024	14/08/2024
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A0709	Device Sensing Problem	14/08/2024	14/08/2024
									A0712	Pacing Problem	14/08/2024	14/08/2024
							F26	No Health Consequences or Impact	A0709	Device Sensing Problem	14/08/2024	14/08/2024
									A0712	Pacing Problem	14/08/2024	14/08/2024
16484	2025	All other reportable incidents	47265	Dual-chamber implantable pacemaker, rate-responsive	CT1685	Cardiac pacemakers	E0601	Arrhythmia	A071102	Failure to Interrogate	07/01/2025	07/01/2025
									A071204	No Pacing	07/01/2025	07/01/2025
							F1901	Additional Surgery	A071102	Failure to Interrogate	07/01/2025	07/01/2025
									A071204	No Pacing	07/01/2025	07/01/2025
					CT257	Implantable cardiac pacemakers	E0601	Arrhythmia	A071102	Failure to Interrogate	07/01/2025	07/01/2025
									A071204	No Pacing	07/01/2025	07/01/2025
							F1901	Additional Surgery	A071102	Failure to Interrogate	07/01/2025	07/01/2025
									A071204	No Pacing	07/01/2025	07/01/2025
					CT2866	Implantable cardiac electrical stimulators	E0601	Arrhythmia	A071102	Failure to Interrogate	07/01/2025	07/01/2025
									A071204	No Pacing	07/01/2025	07/01/2025
							F1901	Additional Surgery	A071102	Failure to Interrogate	07/01/2025	07/01/2025
									A071204	No Pacing	07/01/2025	07/01/2025
16485	2025	All other reportable incidents	35852	Single-chamber implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E210401	Shock from Patient Lead(s)	A0709	Device Sensing Problem		
									A071302	Inappropriate/Inadequate Shock/Stimulation		
							F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A0709	Device Sensing Problem		
									A071302	Inappropriate/Inadequate Shock/Stimulation		
							F1203	Life Threatening Illness or Injury	A0709	Device Sensing Problem		
									A071302	Inappropriate/Inadequate Shock/Stimulation		
16486	2025	All other reportable incidents	47270	Cardiac resynchronization therapy implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	04/02/2025	04/02/2025
									A0908	Incorrect, Inadequate or Imprecise Result or Readings	04/02/2025	04/02/2025
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	04/02/2025	04/02/2025
									A0908	Incorrect, Inadequate or Imprecise Result or Readings	04/02/2025	04/02/2025
							F1903	Device Explantation	A070504	Premature Discharge of Battery	04/02/2025	04/02/2025
									A0908	Incorrect, Inadequate or Imprecise Result or Readings	04/02/2025	04/02/2025
16487	2025	All other reportable incidents	47265	Dual-chamber implantable pacemaker, rate-responsive	CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem		
									A0711	Interrogation Problem		
							F08	Hospitalization or Prolonged Hospitalization	A0705	Battery Problem		
									A0711	Interrogation Problem		
							F1903	Device Explantation	A0705	Battery Problem		
									A0711	Interrogation Problem		
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem		
									A0711	Interrogation Problem		
							F08	Hospitalization or Prolonged Hospitalization	A0705	Battery Problem		
									A0711	Interrogation Problem		
							F1903	Device Explantation	A0705	Battery Problem		
									A0711	Interrogation Problem		
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem		
									A0711	Interrogation Problem		

							F08	Hospitalization or Prolonged Hospitalization	A0705	Battery Problem		
									A0711	Interrogation Problem		
							F1903	Device Explantation	A0705	Battery Problem		
									A0711	Interrogation Problem		
16488	2025	All other reportable incidents	37265	Dual-chamber implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A072201	High Impedance		
							F26	No Health Consequences or Impact	A072201	High Impedance		
16489	2025	All other reportable incidents	47205	Cardiac pulse generator programmer	CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A13	Communication or Transmission Problem		
							F26	No Health Consequences or Impact	A13	Communication or Transmission Problem		
16490	2025	All other reportable incidents	17882	Physiologic-monitoring defibrillation system	CT1684	Cardiac pacemakers and associated devices	E0602	Cardiac Arrest	A0709	Device Sensing Problem		
							F29	Death not Related to Reported Adverse Event	A0709	Device Sensing Problem		
16491	2025	All other reportable incidents	47265	Dual-chamber implantable pacemaker, rate-responsive	CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A0908	Incorrect, Inadequate or Imprecise Result or Readings	03/01/2025	03/01/2025
							F08	Hospitalization or Prolonged Hospitalization	A0908	Incorrect, Inadequate or Imprecise Result or Readings	03/01/2025	03/01/2025
							F1905	Device Revision or Replacement	A0908	Incorrect, Inadequate or Imprecise Result or Readings	03/01/2025	03/01/2025
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A0908	Incorrect, Inadequate or Imprecise Result or Readings	03/01/2025	03/01/2025
							F08	Hospitalization or Prolonged Hospitalization	A0908	Incorrect, Inadequate or Imprecise Result or Readings	03/01/2025	03/01/2025
							F1905	Device Revision or Replacement	A0908	Incorrect, Inadequate or Imprecise Result or Readings	03/01/2025	03/01/2025
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A0908	Incorrect, Inadequate or Imprecise Result or Readings	03/01/2025	03/01/2025
							F08	Hospitalization or Prolonged Hospitalization	A0908	Incorrect, Inadequate or Imprecise Result or Readings	03/01/2025	03/01/2025
							F1905	Device Revision or Replacement	A0908	Incorrect, Inadequate or Imprecise Result or Readings	03/01/2025	03/01/2025
16492	2025		47270	Cardiac resynchronization therapy implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A160503	Inappropriate or Unexpected Reset		
							F26	No Health Consequences or Impact	A160503	Inappropriate or Unexpected Reset		
16493	2025		47265	Dual-chamber implantable pacemaker, rate-responsive	CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem		
							F26	No Health Consequences or Impact	A0705	Battery Problem		
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem		
							F26	No Health Consequences or Impact	A0705	Battery Problem		
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem		
							F26	No Health Consequences or Impact	A0705	Battery Problem		
16494	2025	All other reportable incidents	36106	Implantable pacemaker analyser, intraoperative	CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A0902	Display or Visual Feedback Problem		
									A11	Computer Software Problem		
									A1107	Data Problem		
							F27	Problem Identified During Non-Clinical Procedure	A0902	Display or Visual Feedback Problem		
									A11	Computer Software Problem		
									A1107	Data Problem		
16495	2025		17882	Physiologic-monitoring defibrillation system	CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A1603	Failure of Device to Self-Test		
							F27	Problem Identified During Non-Clinical Procedure	A1603	Failure of Device to Self-Test		
16496	2025	All other reportable incidents	36106	Implantable pacemaker analyser, intraoperative	CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A13	Communication or Transmission Problem		
							F26	No Health Consequences or Impact	A13	Communication or Transmission Problem		
16497	2025	All other reportable incidents	47270	Cardiac resynchronization therapy implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A072201	High Impedance		
									A1301	Failure to Read Input Signal		
							F26	No Health Consequences or Impact	A072201	High Impedance		
									A1301	Failure to Read Input Signal		
16498	2025	All other reportable incidents			CT1684	Cardiac pacemakers and associated devices	E060109	Tachycardia	A070910	Undersensing		
									A0712	Pacing Problem		
									A071202	Inaccurate Synchronization		
							F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A070910	Undersensing		
									A0712	Pacing Problem		
									A071202	Inaccurate Synchronization		
16499	2025	All other reportable incidents	17882	Physiologic-monitoring defibrillation system	CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A0801	Failure to Calibrate		
							F27	Problem Identified During Non-Clinical Procedure	A0801	Failure to Calibrate		
16500	2025	All other reportable incidents	17882	Physiologic-monitoring defibrillation system	CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A0401	Break		
							F26	No Health Consequences or Impact	A0401	Break		
16501	2025	All other reportable incidents	17882	Physiologic-monitoring defibrillation system	CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A0908	Incorrect, Inadequate or Imprecise Result or Readings		
							F27	Problem Identified During Non-Clinical Procedure	A0908	Incorrect, Inadequate or Imprecise Result or Readings		
16502	2025	All other reportable incidents	60789	Intracardiac pacemaker	CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A070102	High Capture Threshold		
							F26	No Health Consequences or Impact	A070102	High Capture Threshold		
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A070102	High Capture Threshold		
							F26	No Health Consequences or Impact	A070102	High Capture Threshold		
					CT2818	Catheterization/Cannulation/Fluid path devices	E2403	No Clinical Signs, Symptoms or Conditions	A070102	High Capture Threshold		
							F26	No Health Consequences or Impact	A070102	High Capture Threshold		
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A070102	High Capture Threshold		
							F26	No Health Consequences or Impact	A070102	High Capture Threshold		
16503	2025	All other reportable incidents	47265	Dual-chamber implantable pacemaker, rate-responsive	CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A0908	Incorrect, Inadequate or Imprecise Result or Readings	04/01/2025	04/01/2025
							F08	Hospitalization or Prolonged Hospitalization	A0908	Incorrect, Inadequate or Imprecise Result or Readings	04/01/2025	04/01/2025
							F1905	Device Revision or Replacement	A0908	Incorrect, Inadequate or Imprecise Result or Readings	04/01/2025	04/01/2025
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A0908	Incorrect, Inadequate or Imprecise Result or Readings	04/01/2025	04/01/2025
							F08	Hospitalization or Prolonged Hospitalization	A0908	Incorrect, Inadequate or Imprecise Result or Readings	04/01/2025	04/01/2025
							F1905	Device Revision or Replacement	A0908	Incorrect, Inadequate or Imprecise Result or Readings	04/01/2025	04/01/2025
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A0908	Incorrect, Inadequate or Imprecise Result or Readings	04/01/2025	04/01/2025
							F08	Hospitalization or Prolonged Hospitalization	A0908	Incorrect, Inadequate or Imprecise Result or Readings	04/01/2025	04/01/2025
							F1905	Device Revision or Replacement	A0908	Incorrect, Inadequate or Imprecise Result or Readings	04/01/2025	04/01/2025
16504	2025	All other reportable incidents	47270	Cardiac resynchronization therapy implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E210401	Shock from Patient Lead(s)	A071302	Inappropriate/Inadequate Shock/Stimulation		
							F08	Hospitalization or Prolonged Hospitalization	A071302	Inappropriate/Inadequate Shock/Stimulation		

							F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A071302	Inappropriate/Inadequate Shock/Stimulation			
16505	2025		47265	Dual-chamber implantable pacemaker, rate-responsive	CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	26/11/2024	26/11/2024
							F1901	Additional Surgery	A24	No Apparent Device Problem	26/11/2024	26/11/2024
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	26/11/2024	26/11/2024
							F1901	Additional Surgery	A24	No Apparent Device Problem	26/11/2024	26/11/2024
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	26/11/2024	26/11/2024
							F1901	Additional Surgery	A24	No Apparent Device Problem	26/11/2024	26/11/2024
16506	2025	All other reportable incidents	35852	Single-chamber implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
16507	2025	All other reportable incidents	47270	Cardiac resynchronization therapy implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A070909	Oversensing		
							F2203	Imaging Required	A070909	Oversensing		
16508	2025	All other reportable incidents	47270	Cardiac resynchronization therapy implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A040101	Fracture		
							F2203	Imaging Required	A040101	Fracture		
16509	2025	All other reportable incidents	35852	Single-chamber implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	10/03/2025	10/03/2025
									A0908	Incorrect, Inadequate or Imprecise Result or Readings	10/03/2025	10/03/2025
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	10/03/2025	10/03/2025
									A0908	Incorrect, Inadequate or Imprecise Result or Readings	10/03/2025	10/03/2025
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	10/03/2025	10/03/2025
									A0908	Incorrect, Inadequate or Imprecise Result or Readings	10/03/2025	10/03/2025
16510	2025	All other reportable incidents	17882	Physiologic-monitoring defibrillation system	CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A071301	Failure to Deliver Shock/Stimulation		
							F26	No Health Consequences or Impact	A071301	Failure to Deliver Shock/Stimulation		
16511	2025	All other reportable incidents	36106	Implantable pacemaker analyser, intraoperative	CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A0902	Display or Visual Feedback Problem		
									A11	Computer Software Problem		
									A1102	Application Program Problem		
							F26	No Health Consequences or Impact	A0902	Display or Visual Feedback Problem		
									A11	Computer Software Problem		
									A1102	Application Program Problem		
16512	2025	All other reportable incidents	36106	Implantable pacemaker analyser, intraoperative	CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A0902	Display or Visual Feedback Problem		
									A11	Computer Software Problem		
									A1102	Application Program Problem		
							F26	No Health Consequences or Impact	A0902	Display or Visual Feedback Problem		
									A11	Computer Software Problem		
									A1102	Application Program Problem		
16513	2025	All other reportable incidents	47270	Cardiac resynchronization therapy implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	29/04/2025	29/04/2025
									A0908	Incorrect, Inadequate or Imprecise Result or Readings	29/04/2025	29/04/2025
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	29/04/2025	29/04/2025
									A0908	Incorrect, Inadequate or Imprecise Result or Readings	29/04/2025	29/04/2025
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	29/04/2025	29/04/2025
									A0908	Incorrect, Inadequate or Imprecise Result or Readings	29/04/2025	29/04/2025
16514	2025	All other reportable incidents	37265	Dual-chamber implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
16515	2025	All other reportable incidents	47267	Single-chamber implantable pacemaker, rate-responsive	CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
16516	2025	All other reportable incidents	35854	Temporary cardiac pacing balloon catheter	CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A0712	Pacing Problem		
							F26	No Health Consequences or Impact	A0712	Pacing Problem		
16517	2025	All other reportable incidents	35852	Single-chamber implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E210401	Shock from Patient Lead(s)	A0705	Battery Problem	01/01/2025	31/01/2025
									A0709	Device Sensing Problem	01/01/2025	31/01/2025
									A071302	Inappropriate/Inadequate Shock/Stimulation	01/01/2025	31/01/2025
							F08	Hospitalization or Prolonged Hospitalization	A0705	Battery Problem	01/01/2025	31/01/2025
									A0709	Device Sensing Problem	01/01/2025	31/01/2025
									A071302	Inappropriate/Inadequate Shock/Stimulation	01/01/2025	31/01/2025
							F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A0705	Battery Problem	01/01/2025	31/01/2025
									A0709	Device Sensing Problem	01/01/2025	31/01/2025
									A071302	Inappropriate/Inadequate Shock/Stimulation	01/01/2025	31/01/2025
							F1203	Life Threatening Illness or Injury	A0705	Battery Problem	01/01/2025	31/01/2025
									A0709	Device Sensing Problem	01/01/2025	31/01/2025
									A071302	Inappropriate/Inadequate Shock/Stimulation	01/01/2025	31/01/2025
							F1905	Device Revision or Replacement	A0705	Battery Problem	01/01/2025	31/01/2025
									A0709	Device Sensing Problem	01/01/2025	31/01/2025
									A071302	Inappropriate/Inadequate Shock/Stimulation	01/01/2025	31/01/2025
16518	2025	All other reportable incidents			CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070909	Oversensing		
									A070910	Undersensing		
									A0712	Pacing Problem		
									A090801	Signal Artifact/Noise		
							F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A070909	Oversensing		
									A070910	Undersensing		
									A0712	Pacing Problem		
									A090801	Signal Artifact/Noise		
16519	2025	All other reportable incidents	17882	Physiologic-monitoring defibrillation system	CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A0801	Failure to Calibrate		
							F27	Problem Identified During Non-Clinical Procedure	A0801	Failure to Calibrate		
16520	2025	All other reportable incidents	60789	Intracardiac pacemaker	CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A051201	Device Dislodged or Dislocated		
									A23	Use of Device Problem		
							F1905	Device Revision or Replacement	A051201	Device Dislodged or Dislocated		
									A23	Use of Device Problem		
							F26	No Health Consequences or Impact	A051201	Device Dislodged or Dislocated		
									A23	Use of Device Problem		
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A051201	Device Dislodged or Dislocated		
									A23	Use of Device Problem		
							F1905	Device Revision or Replacement	A051201	Device Dislodged or Dislocated		

									A23	Use of Device Problem		
							F26	No Health Consequences or Impact	A051201	Device Dislodged or Dislocated		
									A23	Use of Device Problem		
					CT2818	Catheterization/Cannulation/Fluid path devices	E2403	No Clinical Signs, Symptoms or Conditions	A051201	Device Dislodged or Dislocated		
									A23	Use of Device Problem		
							F1905	Device Revision or Replacement	A051201	Device Dislodged or Dislocated		
									A23	Use of Device Problem		
							F26	No Health Consequences or Impact	A051201	Device Dislodged or Dislocated		
									A23	Use of Device Problem		
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A051201	Device Dislodged or Dislocated		
									A23	Use of Device Problem		
							F1905	Device Revision or Replacement	A051201	Device Dislodged or Dislocated		
									A23	Use of Device Problem		
							F26	No Health Consequences or Impact	A051201	Device Dislodged or Dislocated		
									A23	Use of Device Problem		
16521	2025	All other reportable incidents	47270	Cardiac resynchronization therapy implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A070909	Oversensing		
									A072201	High Impedance		
									A1301	Failure to Read Input Signal		
							F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A070909	Oversensing		
									A072201	High Impedance		
									A1301	Failure to Read Input Signal		
16522	2025	All other reportable incidents	17882	Physiologic-monitoring defibrillation system	CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070803	Failure to Power Up		
							F26	No Health Consequences or Impact	A070803	Failure to Power Up		
16523	2025	All other reportable incidents	17882	Physiologic-monitoring defibrillation system	CT1684	Cardiac pacemakers and associated devices	E0601	Arrhythmia	A071301	Failure to Deliver Shock/Stimulation		
							F05	Delay to Treatment/ Therapy	A071301	Failure to Deliver Shock/Stimulation		
16524	2025		47265	Dual-chamber implantable pacemaker, rate-responsive	CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	14/12/2024	14/12/2024
							F1901	Additional Surgery	A24	No Apparent Device Problem	14/12/2024	14/12/2024
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	14/12/2024	14/12/2024
							F1901	Additional Surgery	A24	No Apparent Device Problem	14/12/2024	14/12/2024
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	14/12/2024	14/12/2024
							F1901	Additional Surgery	A24	No Apparent Device Problem	14/12/2024	14/12/2024
16525	2025	All other reportable incidents	47270	Cardiac resynchronization therapy implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A072001	Electromagnetic Interference		
									A160503	Inappropriate or Unexpected Reset		
							F26	No Health Consequences or Impact	A072001	Electromagnetic Interference		
									A160503	Inappropriate or Unexpected Reset		
16526	2025	All other reportable incidents	47940	Triple-chamber implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A070101	Failure to Capture		
							F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A070101	Failure to Capture		
16527	2025	All other reportable incidents	47265	Dual-chamber implantable pacemaker, rate-responsive	CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
16528	2025	All other reportable incidents	17882	Physiologic-monitoring defibrillation system	CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070802	Intermittent Loss of Power		
							F27	Problem Identified During Non-Clinical Procedure	A070802	Intermittent Loss of Power		
16529	2025	All other reportable incidents	47205	Cardiac pulse generator programmer	CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A0902	Display or Visual Feedback Problem		
							F27	Problem Identified During Non-Clinical Procedure	A0902	Display or Visual Feedback Problem		
16530	2025	All other reportable incidents	47267	Single-chamber implantable pacemaker, rate-responsive	CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
16531	2025	All other reportable incidents	37265	Dual-chamber implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A0712	Pacing Problem		
							F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A0712	Pacing Problem		
16532	2025	All other reportable incidents	36106	Implantable pacemaker analyser, intraoperative	CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A11	Computer Software Problem		
							F26	No Health Consequences or Impact	A11	Computer Software Problem		
16533	2025	All other reportable incidents	17882	Physiologic-monitoring defibrillation system	CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A0719	Unexpected Shutdown		
							F26	No Health Consequences or Impact	A0719	Unexpected Shutdown		
16534	2025		47265	Dual-chamber implantable pacemaker, rate-responsive	CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	07/12/2024	07/12/2024
							F1901	Additional Surgery	A24	No Apparent Device Problem	07/12/2024	07/12/2024
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	07/12/2024	07/12/2024
							F1901	Additional Surgery	A24	No Apparent Device Problem	07/12/2024	07/12/2024
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	07/12/2024	07/12/2024
							F1901	Additional Surgery	A24	No Apparent Device Problem	07/12/2024	07/12/2024
16535	2025	All other reportable incidents	47205	Cardiac pulse generator programmer	CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A0902	Display or Visual Feedback Problem		
							F27	Problem Identified During Non-Clinical Procedure	A0902	Display or Visual Feedback Problem		
16536	2025	All other reportable incidents	47263	Cardiac resynchronization therapy implantable pacemaker	CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A0908	Incorrect, Inadequate or Imprecise Result or Readings	09/01/2025	09/01/2025
							F08	Hospitalization or Prolonged Hospitalization	A0908	Incorrect, Inadequate or Imprecise Result or Readings	09/01/2025	09/01/2025
							F1905	Device Revision or Replacement	A0908	Incorrect, Inadequate or Imprecise Result or Readings	09/01/2025	09/01/2025
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A0908	Incorrect, Inadequate or Imprecise Result or Readings	09/01/2025	09/01/2025
							F08	Hospitalization or Prolonged Hospitalization	A0908	Incorrect, Inadequate or Imprecise Result or Readings	09/01/2025	09/01/2025
							F1905	Device Revision or Replacement	A0908	Incorrect, Inadequate or Imprecise Result or Readings	09/01/2025	09/01/2025
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A0908	Incorrect, Inadequate or Imprecise Result or Readings	09/01/2025	09/01/2025
							F08	Hospitalization or Prolonged Hospitalization	A0908	Incorrect, Inadequate or Imprecise Result or Readings	09/01/2025	09/01/2025

							F1905	Device Revision or Replacement	A0908	Incorrect, Inadequate or Imprecise Result or Readings	09/01/2025	09/01/2025
16537	2025	All other reportable incidents	47205	Cardiac pulse generator programmer	CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A13	Communication or Transmission Problem		
							F26	No Health Consequences or Impact	A13	Communication or Transmission Problem		
16538	2025	All other reportable incidents	47205	Cardiac pulse generator programmer	CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A13	Communication or Transmission Problem		
							F26	No Health Consequences or Impact	A13	Communication or Transmission Problem		
16539	2025	All other reportable incidents	37265	Dual-chamber implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E060109	Tachycardia	A0709	Device Sensing Problem		
									A071301	Failure to Deliver Shock/Stimulation		
							F1001	Absence of Treatment	A0709	Device Sensing Problem		
									A071301	Failure to Deliver Shock/Stimulation		
							F1203	Life Threatening Illness or Injury	A0709	Device Sensing Problem		
									A071301	Failure to Deliver Shock/Stimulation		
							F2307	Reprogramming of Device	A0709	Device Sensing Problem		
									A071301	Failure to Deliver Shock/Stimulation		
16540	2025	All other reportable incidents	36106	Implantable pacemaker analyser, intraoperative	CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A09	Output Problem		
									A11	Computer Software Problem		
									A1102	Application Program Problem		
									A1305	Wireless Communication Problem		
							F26	No Health Consequences or Impact	A09	Output Problem		
									A11	Computer Software Problem		
									A1102	Application Program Problem		
									A1305	Wireless Communication Problem		
16541	2025	All other reportable incidents	47270	Cardiac resynchronization therapy implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	13/06/2025	13/06/2025
									A0908	Incorrect, Inadequate or Imprecise Result or Readings	13/06/2025	13/06/2025
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	13/06/2025	13/06/2025
									A0908	Incorrect, Inadequate or Imprecise Result or Readings	13/06/2025	13/06/2025
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	13/06/2025	13/06/2025
									A0908	Incorrect, Inadequate or Imprecise Result or Readings	13/06/2025	13/06/2025
16542	2025	All other reportable incidents	37265	Dual-chamber implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E210401	Shock from Patient Lead(s)	A071201	Failure to Convert Rhythm	28/01/2025	28/01/2025
									A072201	High Impedance	28/01/2025	28/01/2025
									A0908	Incorrect, Inadequate or Imprecise Result or Readings	28/01/2025	28/01/2025
							F08	Hospitalization or Prolonged Hospitalization	A071201	Failure to Convert Rhythm	28/01/2025	28/01/2025
									A072201	High Impedance	28/01/2025	28/01/2025
									A0908	Incorrect, Inadequate or Imprecise Result or Readings	28/01/2025	28/01/2025
							F1905	Device Revision or Replacement	A071201	Failure to Convert Rhythm	28/01/2025	28/01/2025
									A072201	High Impedance	28/01/2025	28/01/2025
									A0908	Incorrect, Inadequate or Imprecise Result or Readings	28/01/2025	28/01/2025
16543	2025	All other reportable incidents	63188	Left-ventricular wireless pacing system sensor/transmitter	CT145	Cardiovascular devices			A070504	Premature Discharge of Battery		
					CT1577	Wireless monitoring system transmitters			A070504	Premature Discharge of Battery		
					CT1684	Cardiac pacemakers and associated devices			A070504	Premature Discharge of Battery		
					CT2727	Active implant devices/ancillaries			A070504	Premature Discharge of Battery		
					CT2750	Implantable cardiac electrical stimulators and associated devices			A070504	Premature Discharge of Battery		
					CT2885	Physiological monitoring/measuring system transmitters/receivers/transceivers			A070504	Premature Discharge of Battery		
16544	2025	All other reportable incidents	47265	Dual-chamber implantable pacemaker, rate-responsive	CT1685	Cardiac pacemakers	E012207	Twitching	A070102	High Capture Threshold		
							F11	Non-Serious Injury/ Illness/ Impairment	A070102	High Capture Threshold		
					CT257	Implantable cardiac pacemakers	E012207	Twitching	A070102	High Capture Threshold		
							F11	Non-Serious Injury/ Illness/ Impairment	A070102	High Capture Threshold		
					CT2866	Implantable cardiac electrical stimulators	E012207	Twitching	A070102	High Capture Threshold		
							F11	Non-Serious Injury/ Illness/ Impairment	A070102	High Capture Threshold		
16545	2025	All other reportable incidents	47265	Dual-chamber implantable pacemaker, rate-responsive	CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A070102	High Capture Threshold		
							F26	No Health Consequences or Impact	A070102	High Capture Threshold		
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A070102	High Capture Threshold		
							F26	No Health Consequences or Impact	A070102	High Capture Threshold		
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A070102	High Capture Threshold		
							F26	No Health Consequences or Impact	A070102	High Capture Threshold		
16546	2025	All other reportable incidents	37265	Dual-chamber implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A070909	Oversensing	16/01/2025	16/01/2025
							F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A070909	Oversensing	16/01/2025	16/01/2025
							F1908	Prolonged Surgery	A070909	Oversensing	16/01/2025	16/01/2025
							F2203	Imaging Required	A070909	Oversensing	16/01/2025	16/01/2025
16547	2025	All other reportable incidents	37265	Dual-chamber implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A070909	Oversensing		
							F05	Delay to Treatment/ Therapy	A070909	Oversensing		
16548	2025	All other reportable incidents	37265	Dual-chamber implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E020201	Anxiety	A071302	Inappropriate/Inadequate Shock/Stimulation		
							E020202	Depression	A071302	Inappropriate/Inadequate Shock/Stimulation		
							E0623	Unspecified Heart Problem	A071302	Inappropriate/Inadequate Shock/Stimulation		
							E210401	Shock from Patient Lead(s)	A071302	Inappropriate/Inadequate Shock/Stimulation		
							E2311	Discomfort	A071302	Inappropriate/Inadequate Shock/Stimulation		
							F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A071302	Inappropriate/Inadequate Shock/Stimulation		
							F15	Recognized Device or Procedural Complication	A071302	Inappropriate/Inadequate Shock/Stimulation		
16549	2025	All other reportable incidents	35852	Single-chamber implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	23/04/2025	23/04/2025
									A0722	Impedance Problem	23/04/2025	23/04/2025
									A0908	Incorrect, Inadequate or Imprecise Result or Readings	23/04/2025	23/04/2025
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	23/04/2025	23/04/2025
									A0722	Impedance Problem	23/04/2025	23/04/2025
									A0908	Incorrect, Inadequate or Imprecise Result or Readings	23/04/2025	23/04/2025
							F1903	Device Explantation	A070504	Premature Discharge of Battery	23/04/2025	23/04/2025
									A0722	Impedance Problem	23/04/2025	23/04/2025
									A0908	Incorrect, Inadequate or Imprecise Result or Readings	23/04/2025	23/04/2025
16550	2025	All other reportable incidents	47265	Dual-chamber implantable pacemaker, rate-responsive	CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A05	Mechanical Problem	23/01/2025	23/01/2025
							F26	No Health Consequences or Impact	A05	Mechanical Problem	23/01/2025	23/01/2025
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A05	Mechanical Problem	23/01/2025	23/01/2025

							F26	No Health Consequences or Impact	A05	Mechanical Problem	23/01/2025	23/01/2025
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A05	Mechanical Problem	23/01/2025	23/01/2025
							F26	No Health Consequences or Impact	A05	Mechanical Problem	23/01/2025	23/01/2025
16551	2025	All other reportable incidents	47265	Dual-chamber implantable pacemaker, rate-responsive	CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	20/11/2023	20/11/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	20/11/2023	20/11/2023
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	20/11/2023	20/11/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	20/11/2023	20/11/2023
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	20/11/2023	20/11/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	20/11/2023	20/11/2023
16552	2025	All other reportable incidents	47267	Single-chamber implantable pacemaker, rate-responsive	CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	01/01/2025	31/12/2025
							F12	Serious Injury/ Illness/ Impairment	A070504	Premature Discharge of Battery	01/01/2025	31/12/2025
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	01/01/2025	31/12/2025
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery	01/01/2025	31/12/2025
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	01/01/2025	31/12/2025
							F12	Serious Injury/ Illness/ Impairment	A070504	Premature Discharge of Battery	01/01/2025	31/12/2025
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	01/01/2025	31/12/2025
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery	01/01/2025	31/12/2025
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	01/01/2025	31/12/2025
							F12	Serious Injury/ Illness/ Impairment	A070504	Premature Discharge of Battery	01/01/2025	31/12/2025
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	01/01/2025	31/12/2025
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery	01/01/2025	31/12/2025
16553	2025	All other reportable incidents	47263	Cardiac resynchronization therapy implantable pacemaker	CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A0908	Incorrect, Inadequate or Imprecise Result or Readings	20/01/2025	20/01/2025
							F08	Hospitalization or Prolonged Hospitalization	A0908	Incorrect, Inadequate or Imprecise Result or Readings	20/01/2025	20/01/2025
							F1905	Device Revision or Replacement	A0908	Incorrect, Inadequate or Imprecise Result or Readings	20/01/2025	20/01/2025
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A0908	Incorrect, Inadequate or Imprecise Result or Readings	20/01/2025	20/01/2025
							F08	Hospitalization or Prolonged Hospitalization	A0908	Incorrect, Inadequate or Imprecise Result or Readings	20/01/2025	20/01/2025
							F1905	Device Revision or Replacement	A0908	Incorrect, Inadequate or Imprecise Result or Readings	20/01/2025	20/01/2025
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A0908	Incorrect, Inadequate or Imprecise Result or Readings	20/01/2025	20/01/2025
							F08	Hospitalization or Prolonged Hospitalization	A0908	Incorrect, Inadequate or Imprecise Result or Readings	20/01/2025	20/01/2025
							F1905	Device Revision or Replacement	A0908	Incorrect, Inadequate or Imprecise Result or Readings	20/01/2025	20/01/2025
16554	2025	All other reportable incidents	47263	Cardiac resynchronization therapy implantable pacemaker	CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A070910	Undersensing		
									A1301	Failure to Read Input Signal		
							F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A070910	Undersensing		
									A1301	Failure to Read Input Signal		
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A070910	Undersensing		
									A1301	Failure to Read Input Signal		
							F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A070910	Undersensing		
									A1301	Failure to Read Input Signal		
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A070910	Undersensing		
									A1301	Failure to Read Input Signal		
							F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A070910	Undersensing		
									A1301	Failure to Read Input Signal		
16555	2025	All other reportable incidents	47270	Cardiac resynchronization therapy implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	27/03/2025	27/03/2025
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery	27/03/2025	27/03/2025
16556	2025	All other reportable incidents	47267	Single-chamber implantable pacemaker, rate-responsive	CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A0701	Capturing Problem		
									A070101	Failure to Capture		
									A24	No Apparent Device Problem		
							F12	Serious Injury/ Illness/ Impairment	A0701	Capturing Problem		
									A070101	Failure to Capture		
									A24	No Apparent Device Problem		
							F2307	Reprogramming of Device	A0701	Capturing Problem		
									A070101	Failure to Capture		
									A24	No Apparent Device Problem		
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A0701	Capturing Problem		
									A070101	Failure to Capture		
									A24	No Apparent Device Problem		
							F12	Serious Injury/ Illness/ Impairment	A0701	Capturing Problem		
									A070101	Failure to Capture		
									A24	No Apparent Device Problem		
							F2307	Reprogramming of Device	A0701	Capturing Problem		
									A070101	Failure to Capture		
									A24	No Apparent Device Problem		
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A0701	Capturing Problem		
									A070101	Failure to Capture		
									A24	No Apparent Device Problem		
							F12	Serious Injury/ Illness/ Impairment	A0701	Capturing Problem		
									A070101	Failure to Capture		
									A24	No Apparent Device Problem		
							F2307	Reprogramming of Device	A0701	Capturing Problem		
									A070101	Failure to Capture		
									A24	No Apparent Device Problem		
16557	2025		60956	Pacing/defibrillation lead tunneller	CT1684	Cardiac pacemakers and associated devices	E2104	Electric Shock	A04	Material Integrity Problem		
									A071302	Inappropriate/Inadequate Shock/Stimulation		
							F15	Recognized Device or Procedural Complication	A04	Material Integrity Problem		
									A071302	Inappropriate/Inadequate Shock/Stimulation		
							F1903	Device Explantation	A04	Material Integrity Problem		
									A071302	Inappropriate/Inadequate Shock/Stimulation		
16558	2025	All other reportable incidents	47270	Cardiac resynchronization therapy implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	19/03/2025	19/03/2025
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	19/03/2025	19/03/2025
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	19/03/2025	19/03/2025
16559	2025	All other reportable incidents	47267	Single-chamber implantable pacemaker, rate-			E0112	Dizziness	A0712	Pacing Problem	17/01/2025	17/01/2025

				responsive					A0908	Incorrect, Inadequate or Imprecise Result or Readings	17/01/2025	17/01/2025
							E011903	Syncope/Fainting	A0712	Pacing Problem	17/01/2025	17/01/2025
									A0908	Incorrect, Inadequate or Imprecise Result or Readings	17/01/2025	17/01/2025
							E060101	Asystole	A0712	Pacing Problem	17/01/2025	17/01/2025
									A0908	Incorrect, Inadequate or Imprecise Result or Readings	17/01/2025	17/01/2025
							F08	Hospitalization or Prolonged Hospitalization	A0712	Pacing Problem	17/01/2025	17/01/2025
									A0908	Incorrect, Inadequate or Imprecise Result or Readings	17/01/2025	17/01/2025
							F1001	Absence of Treatment	A0712	Pacing Problem	17/01/2025	17/01/2025
									A0908	Incorrect, Inadequate or Imprecise Result or Readings	17/01/2025	17/01/2025
							F1905	Device Revision or Replacement	A0712	Pacing Problem	17/01/2025	17/01/2025
									A0908	Incorrect, Inadequate or Imprecise Result or Readings	17/01/2025	17/01/2025
16560	2025	All other reportable incidents	47265	Dual-chamber implantable pacemaker, rate-responsive	CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	24/01/2025	24/01/2025
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	24/01/2025	24/01/2025
							F12	Serious Injury/ Illness/ Impairment	A070504	Premature Discharge of Battery	24/01/2025	24/01/2025
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	24/01/2025	24/01/2025
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	24/01/2025	24/01/2025
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	24/01/2025	24/01/2025
							F12	Serious Injury/ Illness/ Impairment	A070504	Premature Discharge of Battery	24/01/2025	24/01/2025
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	24/01/2025	24/01/2025
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	24/01/2025	24/01/2025
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	24/01/2025	24/01/2025
							F12	Serious Injury/ Illness/ Impairment	A070504	Premature Discharge of Battery	24/01/2025	24/01/2025
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	24/01/2025	24/01/2025
16561	2025	All other reportable incidents	47265	Dual-chamber implantable pacemaker, rate-responsive	CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A160105	False Alarm		
							F26	No Health Consequences or Impact	A160105	False Alarm		
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A160105	False Alarm		
							F26	No Health Consequences or Impact	A160105	False Alarm		
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A160105	False Alarm		
							F26	No Health Consequences or Impact	A160105	False Alarm		
16562	2025	All other reportable incidents	17882	Physiologic-monitoring defibrillation system	CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A0507	Mechanics Altered		
							F27	Problem Identified During Non-Clinical Procedure	A0507	Mechanics Altered		
16563	2025	All other reportable incidents	47270	Cardiac resynchronization therapy implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	08/05/2025	08/05/2025
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	08/05/2025	08/05/2025
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	08/05/2025	08/05/2025
16564	2025	All other reportable incidents	37265	Dual-chamber implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
									A070602	Delayed Charge Time		
							F26	No Health Consequences or Impact	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
									A070602	Delayed Charge Time		
16565	2025	All other reportable incidents	47267	Single-chamber implantable pacemaker, rate-responsive			E2403	No Clinical Signs, Symptoms or Conditions	A0908	Incorrect, Inadequate or Imprecise Result or Readings	14/01/2025	14/01/2025
							F08	Hospitalization or Prolonged Hospitalization	A0908	Incorrect, Inadequate or Imprecise Result or Readings	14/01/2025	14/01/2025
							F1905	Device Revision or Replacement	A0908	Incorrect, Inadequate or Imprecise Result or Readings	14/01/2025	14/01/2025
16566	2025	All other reportable incidents	47267	Single-chamber implantable pacemaker, rate-responsive	CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A05	Mechanical Problem	25/01/2025	25/01/2025
							F26	No Health Consequences or Impact	A05	Mechanical Problem	25/01/2025	25/01/2025
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A05	Mechanical Problem	25/01/2025	25/01/2025
							F26	No Health Consequences or Impact	A05	Mechanical Problem	25/01/2025	25/01/2025
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A05	Mechanical Problem	25/01/2025	25/01/2025
							F26	No Health Consequences or Impact	A05	Mechanical Problem	25/01/2025	25/01/2025
16567	2025	All other reportable incidents	47270	Cardiac resynchronization therapy implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A0508	Noise, Audible		
									A070909	Oversensing		
									A072201	High Impedance		
									A090801	Signal Artifact/Noise		
									A1301	Failure to Read Input Signal		
							F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A0508	Noise, Audible		
									A070909	Oversensing		
									A072201	High Impedance		
									A090801	Signal Artifact/Noise		
									A1301	Failure to Read Input Signal		
							F2203	Imaging Required	A0508	Noise, Audible		
									A070909	Oversensing		
									A072201	High Impedance		
									A090801	Signal Artifact/Noise		
									A1301	Failure to Read Input Signal		
16568	2025	All other reportable incidents	37265	Dual-chamber implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E060110	Ventricular Fibrillation	A070910	Undersensing		
									A0712	Pacing Problem		
									A071202	Inaccurate Synchronization		
							F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A070910	Undersensing		
									A0712	Pacing Problem		
									A071202	Inaccurate Synchronization		
							F1203	Life Threatening Illness or Injury	A070910	Undersensing		
									A0712	Pacing Problem		
									A071202	Inaccurate Synchronization		
							F15	Recognized Device or Procedural Complication	A070910	Undersensing		
									A0712	Pacing Problem		
									A071202	Inaccurate Synchronization		
16569	2025	All other reportable incidents	47205	Cardiac pulse generator programmer	CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A0902	Display or Visual Feedback Problem		
							F27	Problem Identified During Non-Clinical Procedure	A0902	Display or Visual Feedback Problem		
16570	2025	All other reportable incidents	47265	Dual-chamber implantable pacemaker, rate-responsive	CT1685	Cardiac pacemakers	E1708	Itching Sensation	A27	Appropriate Device Problem Term/Code not Available		

							E1714	Rash	A27	Appropriate Device Problem Term/Code not Available		
							F08	Hospitalization or Prolonged Hospitalization	A27	Appropriate Device Problem Term/Code not Available		
							F12	Serious Injury/ Illness/ Impairment	A27	Appropriate Device Problem Term/Code not Available		
					CT257	Implantable cardiac pacemakers	E1708	Itching Sensation	A27	Appropriate Device Problem Term/Code not Available		
							E1714	Rash	A27	Appropriate Device Problem Term/Code not Available		
							F08	Hospitalization or Prolonged Hospitalization	A27	Appropriate Device Problem Term/Code not Available		
							F12	Serious Injury/ Illness/ Impairment	A27	Appropriate Device Problem Term/Code not Available		
					CT2866	Implantable cardiac electrical stimulators	E1708	Itching Sensation	A27	Appropriate Device Problem Term/Code not Available		
							E1714	Rash	A27	Appropriate Device Problem Term/Code not Available		
							F08	Hospitalization or Prolonged Hospitalization	A27	Appropriate Device Problem Term/Code not Available		
							F12	Serious Injury/ Illness/ Impairment	A27	Appropriate Device Problem Term/Code not Available		
16571	2025	All other reportable incidents	47265	Dual-chamber implantable pacemaker, rate-responsive	CT1685	Cardiac pacemakers	E0112	Dizziness	A070909	Oversensing	27/01/2025	27/01/2025
									A0712	Pacing Problem	27/01/2025	27/01/2025
									A0908	Incorrect, Inadequate or Imprecise Result or Readings	27/01/2025	27/01/2025
									A090801	Signal Artifact/Noise	27/01/2025	27/01/2025
							E0119	Loss of Consciousness	A070909	Oversensing	27/01/2025	27/01/2025
									A0712	Pacing Problem	27/01/2025	27/01/2025
									A0908	Incorrect, Inadequate or Imprecise Result or Readings	27/01/2025	27/01/2025
									A090801	Signal Artifact/Noise	27/01/2025	27/01/2025
							E011903	Syncope/Fainting	A070909	Oversensing	27/01/2025	27/01/2025
									A0712	Pacing Problem	27/01/2025	27/01/2025
									A0908	Incorrect, Inadequate or Imprecise Result or Readings	27/01/2025	27/01/2025
									A090801	Signal Artifact/Noise	27/01/2025	27/01/2025
							F08	Hospitalization or Prolonged Hospitalization	A070909	Oversensing	27/01/2025	27/01/2025
									A0712	Pacing Problem	27/01/2025	27/01/2025
									A0908	Incorrect, Inadequate or Imprecise Result or Readings	27/01/2025	27/01/2025
									A090801	Signal Artifact/Noise	27/01/2025	27/01/2025
							F1001	Absence of Treatment	A070909	Oversensing	27/01/2025	27/01/2025
									A0712	Pacing Problem	27/01/2025	27/01/2025
									A0908	Incorrect, Inadequate or Imprecise Result or Readings	27/01/2025	27/01/2025
									A090801	Signal Artifact/Noise	27/01/2025	27/01/2025
							F1905	Device Revision or Replacement	A070909	Oversensing	27/01/2025	27/01/2025
									A0712	Pacing Problem	27/01/2025	27/01/2025
									A0908	Incorrect, Inadequate or Imprecise Result or Readings	27/01/2025	27/01/2025
									A090801	Signal Artifact/Noise	27/01/2025	27/01/2025
					CT257	Implantable cardiac pacemakers	E0112	Dizziness	A070909	Oversensing	27/01/2025	27/01/2025
									A0712	Pacing Problem	27/01/2025	27/01/2025
									A0908	Incorrect, Inadequate or Imprecise Result or Readings	27/01/2025	27/01/2025
									A090801	Signal Artifact/Noise	27/01/2025	27/01/2025
							E0119	Loss of Consciousness	A070909	Oversensing	27/01/2025	27/01/2025
									A0712	Pacing Problem	27/01/2025	27/01/2025
									A0908	Incorrect, Inadequate or Imprecise Result or Readings	27/01/2025	27/01/2025
									A090801	Signal Artifact/Noise	27/01/2025	27/01/2025
							E011903	Syncope/Fainting	A070909	Oversensing	27/01/2025	27/01/2025
									A0712	Pacing Problem	27/01/2025	27/01/2025
									A0908	Incorrect, Inadequate or Imprecise Result or Readings	27/01/2025	27/01/2025
									A090801	Signal Artifact/Noise	27/01/2025	27/01/2025
							F08	Hospitalization or Prolonged Hospitalization	A070909	Oversensing	27/01/2025	27/01/2025
									A0712	Pacing Problem	27/01/2025	27/01/2025
									A0908	Incorrect, Inadequate or Imprecise Result or Readings	27/01/2025	27/01/2025
									A090801	Signal Artifact/Noise	27/01/2025	27/01/2025
							F1001	Absence of Treatment	A070909	Oversensing	27/01/2025	27/01/2025
									A0712	Pacing Problem	27/01/2025	27/01/2025
									A0908	Incorrect, Inadequate or Imprecise Result or Readings	27/01/2025	27/01/2025
									A090801	Signal Artifact/Noise	27/01/2025	27/01/2025
							F1905	Device Revision or Replacement	A070909	Oversensing	27/01/2025	27/01/2025
									A0712	Pacing Problem	27/01/2025	27/01/2025
									A0908	Incorrect, Inadequate or Imprecise Result or Readings	27/01/2025	27/01/2025
									A090801	Signal Artifact/Noise	27/01/2025	27/01/2025
					CT2866	Implantable cardiac electrical stimulators	E0112	Dizziness	A070909	Oversensing	27/01/2025	27/01/2025
									A0712	Pacing Problem	27/01/2025	27/01/2025
									A0908	Incorrect, Inadequate or Imprecise Result or Readings	27/01/2025	27/01/2025
									A090801	Signal Artifact/Noise	27/01/2025	27/01/2025
							E0119	Loss of Consciousness	A070909	Oversensing	27/01/2025	27/01/2025
									A0712	Pacing Problem	27/01/2025	27/01/2025
									A0908	Incorrect, Inadequate or Imprecise Result or Readings	27/01/2025	27/01/2025
									A090801	Signal Artifact/Noise	27/01/2025	27/01/2025
							E011903	Syncope/Fainting	A070909	Oversensing	27/01/2025	27/01/2025
									A0712	Pacing Problem	27/01/2025	27/01/2025

									A0908	Incorrect, Inadequate or Imprecise Result or Readings	27/01/2025	27/01/2025
									A090801	Signal Artifact/Noise	27/01/2025	27/01/2025
							F08	Hospitalization or Prolonged Hospitalization	A070909	Oversensing	27/01/2025	27/01/2025
									A0712	Pacing Problem	27/01/2025	27/01/2025
									A0908	Incorrect, Inadequate or Imprecise Result or Readings	27/01/2025	27/01/2025
									A090801	Signal Artifact/Noise	27/01/2025	27/01/2025
							F1001	Absence of Treatment	A070909	Oversensing	27/01/2025	27/01/2025
									A0712	Pacing Problem	27/01/2025	27/01/2025
									A0908	Incorrect, Inadequate or Imprecise Result or Readings	27/01/2025	27/01/2025
									A090801	Signal Artifact/Noise	27/01/2025	27/01/2025
							F1905	Device Revision or Replacement	A070909	Oversensing	27/01/2025	27/01/2025
									A0712	Pacing Problem	27/01/2025	27/01/2025
									A0908	Incorrect, Inadequate or Imprecise Result or Readings	27/01/2025	27/01/2025
									A090801	Signal Artifact/Noise	27/01/2025	27/01/2025
16572	2025	All other reportable incidents	35852	Single-chamber implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem	28/01/2025	28/01/2025
									A070504	Premature Discharge of Battery	28/01/2025	28/01/2025
									A1605	Reset Problem	28/01/2025	28/01/2025
							F08	Hospitalization or Prolonged Hospitalization	A0705	Battery Problem	28/01/2025	28/01/2025
									A070504	Premature Discharge of Battery	28/01/2025	28/01/2025
									A1605	Reset Problem	28/01/2025	28/01/2025
							F12	Serious Injury/ Illness/ Impairment	A0705	Battery Problem	28/01/2025	28/01/2025
									A070504	Premature Discharge of Battery	28/01/2025	28/01/2025
									A1605	Reset Problem	28/01/2025	28/01/2025
							F1905	Device Revision or Replacement	A0705	Battery Problem	28/01/2025	28/01/2025
									A070504	Premature Discharge of Battery	28/01/2025	28/01/2025
									A1605	Reset Problem	28/01/2025	28/01/2025
							F2307	Reprogramming of Device	A0705	Battery Problem	28/01/2025	28/01/2025
									A070504	Premature Discharge of Battery	28/01/2025	28/01/2025
									A1605	Reset Problem	28/01/2025	28/01/2025
16573	2025	All other reportable incidents	47270	Cardiac resynchronization therapy implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E210401	Shock from Patient Lead(s)	A070909	Oversensing	24/01/2025	24/01/2025
									A071302	Inappropriate/Inadequate Shock/Stimulation	24/01/2025	24/01/2025
							F08	Hospitalization or Prolonged Hospitalization	A070909	Oversensing	24/01/2025	24/01/2025
									A071302	Inappropriate/Inadequate Shock/Stimulation	24/01/2025	24/01/2025
							F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A070909	Oversensing	24/01/2025	24/01/2025
									A071302	Inappropriate/Inadequate Shock/Stimulation	24/01/2025	24/01/2025
							F1903	Device Explantation	A070909	Oversensing	24/01/2025	24/01/2025
									A071302	Inappropriate/Inadequate Shock/Stimulation	24/01/2025	24/01/2025
							F23	Unexpected Medical Intervention	A070909	Oversensing	24/01/2025	24/01/2025
									A071302	Inappropriate/Inadequate Shock/Stimulation	24/01/2025	24/01/2025
16574	2025	All other reportable incidents	17882	Physiologic-monitoring defibrillation system	CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A1603	Failure of Device to Self-Test		
							F27	Problem Identified During Non-Clinical Procedure	A1603	Failure of Device to Self-Test		
16575	2025	All other reportable incidents	35852	Single-chamber implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E210401	Shock from Patient Lead(s)	A0712	Pacing Problem		
									A071302	Inappropriate/Inadequate Shock/Stimulation		
							F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A0712	Pacing Problem		
									A071302	Inappropriate/Inadequate Shock/Stimulation		
16576	2025	All other reportable incidents	47265	Dual-chamber implantable pacemaker, rate-responsive	CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A0722	Impedance Problem	28/01/2025	28/01/2025
									A0908	Incorrect, Inadequate or Imprecise Result or Readings	28/01/2025	28/01/2025
							F08	Hospitalization or Prolonged Hospitalization	A0722	Impedance Problem	28/01/2025	28/01/2025
									A0908	Incorrect, Inadequate or Imprecise Result or Readings	28/01/2025	28/01/2025
							F1905	Device Revision or Replacement	A0722	Impedance Problem	28/01/2025	28/01/2025
									A0908	Incorrect, Inadequate or Imprecise Result or Readings	28/01/2025	28/01/2025
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A0722	Impedance Problem	28/01/2025	28/01/2025
									A0908	Incorrect, Inadequate or Imprecise Result or Readings	28/01/2025	28/01/2025
							F08	Hospitalization or Prolonged Hospitalization	A0722	Impedance Problem	28/01/2025	28/01/2025
									A0908	Incorrect, Inadequate or Imprecise Result or Readings	28/01/2025	28/01/2025
							F1905	Device Revision or Replacement	A0722	Impedance Problem	28/01/2025	28/01/2025
									A0908	Incorrect, Inadequate or Imprecise Result or Readings	28/01/2025	28/01/2025
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A0722	Impedance Problem	28/01/2025	28/01/2025
									A0908	Incorrect, Inadequate or Imprecise Result or Readings	28/01/2025	28/01/2025
							F08	Hospitalization or Prolonged Hospitalization	A0722	Impedance Problem	28/01/2025	28/01/2025
									A0908	Incorrect, Inadequate or Imprecise Result or Readings	28/01/2025	28/01/2025
							F1905	Device Revision or Replacement	A0722	Impedance Problem	28/01/2025	28/01/2025
									A0908	Incorrect, Inadequate or Imprecise Result or Readings	28/01/2025	28/01/2025
16577	2025	All other reportable incidents	47267	Single-chamber implantable pacemaker, rate-responsive	CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A05	Mechanical Problem		
							F26	No Health Consequences or Impact	A05	Mechanical Problem		
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A05	Mechanical Problem		
							F26	No Health Consequences or Impact	A05	Mechanical Problem		
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A05	Mechanical Problem		
							F26	No Health Consequences or Impact	A05	Mechanical Problem		
16578	2025	All other reportable incidents	47265	Dual-chamber implantable pacemaker, rate-responsive	CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem	14/01/2025	14/01/2025
									A070504	Premature Discharge of Battery	14/01/2025	14/01/2025
									A0908	Incorrect, Inadequate or Imprecise Result or Readings	14/01/2025	14/01/2025
							F08	Hospitalization or Prolonged Hospitalization	A0705	Battery Problem	14/01/2025	14/01/2025
									A070504	Premature Discharge of Battery	14/01/2025	14/01/2025
									A0908	Incorrect, Inadequate or Imprecise Result or Readings	14/01/2025	14/01/2025
							F1905	Device Revision or Replacement	A0705	Battery Problem	14/01/2025	14/01/2025

								A070504	Premature Discharge of Battery	14/01/2025	14/01/2025
								A0908	Incorrect, Inadequate or Imprecise Result or Readings	14/01/2025	14/01/2025
				CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem	14/01/2025	14/01/2025
								A070504	Premature Discharge of Battery	14/01/2025	14/01/2025
								A0908	Incorrect, Inadequate or Imprecise Result or Readings	14/01/2025	14/01/2025
						F08	Hospitalization or Prolonged Hospitalization	A0705	Battery Problem	14/01/2025	14/01/2025
								A070504	Premature Discharge of Battery	14/01/2025	14/01/2025
								A0908	Incorrect, Inadequate or Imprecise Result or Readings	14/01/2025	14/01/2025
						F1905	Device Revision or Replacement	A0705	Battery Problem	14/01/2025	14/01/2025
								A070504	Premature Discharge of Battery	14/01/2025	14/01/2025
								A0908	Incorrect, Inadequate or Imprecise Result or Readings	14/01/2025	14/01/2025
				CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem	14/01/2025	14/01/2025
								A070504	Premature Discharge of Battery	14/01/2025	14/01/2025
								A0908	Incorrect, Inadequate or Imprecise Result or Readings	14/01/2025	14/01/2025
						F08	Hospitalization or Prolonged Hospitalization	A0705	Battery Problem	14/01/2025	14/01/2025
								A070504	Premature Discharge of Battery	14/01/2025	14/01/2025
								A0908	Incorrect, Inadequate or Imprecise Result or Readings	14/01/2025	14/01/2025
						F1905	Device Revision or Replacement	A0705	Battery Problem	14/01/2025	14/01/2025
								A070504	Premature Discharge of Battery	14/01/2025	14/01/2025
								A0908	Incorrect, Inadequate or Imprecise Result or Readings	14/01/2025	14/01/2025
16579	2025	All other reportable incidents	47265	Dual-chamber implantable pacemaker, rate-responsive	CT1685	Cardiac pacemakers	E060104	Bradycardia	A0712	Pacing Problem	04/02/2025	04/02/2025
									A0908	Incorrect, Inadequate or Imprecise Result or Readings	04/02/2025	04/02/2025
						F08	Hospitalization or Prolonged Hospitalization	A0712	Pacing Problem	04/02/2025	04/02/2025
								A0908	Incorrect, Inadequate or Imprecise Result or Readings	04/02/2025	04/02/2025
						F1001	Absence of Treatment	A0712	Pacing Problem	04/02/2025	04/02/2025
								A0908	Incorrect, Inadequate or Imprecise Result or Readings	04/02/2025	04/02/2025
						F19	Surgical Intervention	A0712	Pacing Problem	04/02/2025	04/02/2025
								A0908	Incorrect, Inadequate or Imprecise Result or Readings	04/02/2025	04/02/2025
						F1905	Device Revision or Replacement	A0712	Pacing Problem	04/02/2025	04/02/2025
								A0908	Incorrect, Inadequate or Imprecise Result or Readings	04/02/2025	04/02/2025
				CT257	Implantable cardiac pacemakers	E060104	Bradycardia	A0712	Pacing Problem	04/02/2025	04/02/2025
								A0908	Incorrect, Inadequate or Imprecise Result or Readings	04/02/2025	04/02/2025
						F08	Hospitalization or Prolonged Hospitalization	A0712	Pacing Problem	04/02/2025	04/02/2025
								A0908	Incorrect, Inadequate or Imprecise Result or Readings	04/02/2025	04/02/2025
						F1001	Absence of Treatment	A0712	Pacing Problem	04/02/2025	04/02/2025
								A0908	Incorrect, Inadequate or Imprecise Result or Readings	04/02/2025	04/02/2025
						F19	Surgical Intervention	A0712	Pacing Problem	04/02/2025	04/02/2025
								A0908	Incorrect, Inadequate or Imprecise Result or Readings	04/02/2025	04/02/2025
						F1905	Device Revision or Replacement	A0712	Pacing Problem	04/02/2025	04/02/2025
								A0908	Incorrect, Inadequate or Imprecise Result or Readings	04/02/2025	04/02/2025
				CT2866	Implantable cardiac electrical stimulators	E060104	Bradycardia	A0712	Pacing Problem	04/02/2025	04/02/2025
								A0908	Incorrect, Inadequate or Imprecise Result or Readings	04/02/2025	04/02/2025
						F08	Hospitalization or Prolonged Hospitalization	A0712	Pacing Problem	04/02/2025	04/02/2025
								A0908	Incorrect, Inadequate or Imprecise Result or Readings	04/02/2025	04/02/2025
						F1001	Absence of Treatment	A0712	Pacing Problem	04/02/2025	04/02/2025
								A0908	Incorrect, Inadequate or Imprecise Result or Readings	04/02/2025	04/02/2025
						F19	Surgical Intervention	A0712	Pacing Problem	04/02/2025	04/02/2025
								A0908	Incorrect, Inadequate or Imprecise Result or Readings	04/02/2025	04/02/2025
						F1905	Device Revision or Replacement	A0712	Pacing Problem	04/02/2025	04/02/2025
								A0908	Incorrect, Inadequate or Imprecise Result or Readings	04/02/2025	04/02/2025
16580	2025	All other reportable incidents	36106	Implantable pacemaker analyser, intraoperative	CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A11	Computer Software Problem		
									A1107	Data Problem		
									A13	Communication or Transmission Problem		
						F27	Problem Identified During Non-Clinical Procedure	A11	Computer Software Problem		
								A1107	Data Problem		
								A13	Communication or Transmission Problem		
16581	2025	All other reportable incidents	37265	Dual-chamber implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A072001	Electromagnetic Interference		
							F26	No Health Consequences or Impact	A072001	Electromagnetic Interference		
16582	2025	All other reportable incidents	47265	Dual-chamber implantable pacemaker, rate-responsive	CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	10/01/2025	10/01/2025
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	10/01/2025	10/01/2025
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	10/01/2025	10/01/2025
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	10/01/2025	10/01/2025
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	10/01/2025	10/01/2025
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	10/01/2025	10/01/2025
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	10/01/2025	10/01/2025
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	10/01/2025	10/01/2025
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	10/01/2025	10/01/2025
16583	2025	All other reportable incidents	36106	Implantable pacemaker analyser, intraoperative	CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A0902	Display or Visual Feedback Problem		
									A11	Computer Software Problem		
									A1107	Data Problem		
						F27	Problem Identified During Non-Clinical Procedure	A0902	Display or Visual Feedback Problem		

									A11	Computer Software Problem		
									A1107	Data Problem		
16584	2025	All other reportable incidents	47265	Dual-chamber implantable pacemaker, rate-responsive	CT1685	Cardiac pacemakers	E2103	Device Overstimulation of Tissue	A070504	Premature Discharge of Battery	03/02/2025	03/02/2025
									A0711	Interrogation Problem	03/02/2025	03/02/2025
									A0908	Incorrect, Inadequate or Imprecise Result or Readings	03/02/2025	03/02/2025
							E2311	Discomfort	A070504	Premature Discharge of Battery	03/02/2025	03/02/2025
									A0711	Interrogation Problem	03/02/2025	03/02/2025
									A0908	Incorrect, Inadequate or Imprecise Result or Readings	03/02/2025	03/02/2025
							E2330	Pain	A070504	Premature Discharge of Battery	03/02/2025	03/02/2025
									A0711	Interrogation Problem	03/02/2025	03/02/2025
									A0908	Incorrect, Inadequate or Imprecise Result or Readings	03/02/2025	03/02/2025
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	03/02/2025	03/02/2025
									A0711	Interrogation Problem	03/02/2025	03/02/2025
									A0908	Incorrect, Inadequate or Imprecise Result or Readings	03/02/2025	03/02/2025
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	03/02/2025	03/02/2025
									A0711	Interrogation Problem	03/02/2025	03/02/2025
									A0908	Incorrect, Inadequate or Imprecise Result or Readings	03/02/2025	03/02/2025
					CT257	Implantable cardiac pacemakers	E2103	Device Overstimulation of Tissue	A070504	Premature Discharge of Battery	03/02/2025	03/02/2025
									A0711	Interrogation Problem	03/02/2025	03/02/2025
									A0908	Incorrect, Inadequate or Imprecise Result or Readings	03/02/2025	03/02/2025
							E2311	Discomfort	A070504	Premature Discharge of Battery	03/02/2025	03/02/2025
									A0711	Interrogation Problem	03/02/2025	03/02/2025
									A0908	Incorrect, Inadequate or Imprecise Result or Readings	03/02/2025	03/02/2025
							E2330	Pain	A070504	Premature Discharge of Battery	03/02/2025	03/02/2025
									A0711	Interrogation Problem	03/02/2025	03/02/2025
									A0908	Incorrect, Inadequate or Imprecise Result or Readings	03/02/2025	03/02/2025
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	03/02/2025	03/02/2025
									A0711	Interrogation Problem	03/02/2025	03/02/2025
									A0908	Incorrect, Inadequate or Imprecise Result or Readings	03/02/2025	03/02/2025
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	03/02/2025	03/02/2025
									A0711	Interrogation Problem	03/02/2025	03/02/2025
									A0908	Incorrect, Inadequate or Imprecise Result or Readings	03/02/2025	03/02/2025
					CT2866	Implantable cardiac electrical stimulators	E2103	Device Overstimulation of Tissue	A070504	Premature Discharge of Battery	03/02/2025	03/02/2025
									A0711	Interrogation Problem	03/02/2025	03/02/2025
									A0908	Incorrect, Inadequate or Imprecise Result or Readings	03/02/2025	03/02/2025
							E2311	Discomfort	A070504	Premature Discharge of Battery	03/02/2025	03/02/2025
									A0711	Interrogation Problem	03/02/2025	03/02/2025
									A0908	Incorrect, Inadequate or Imprecise Result or Readings	03/02/2025	03/02/2025
							E2330	Pain	A070504	Premature Discharge of Battery	03/02/2025	03/02/2025
									A0711	Interrogation Problem	03/02/2025	03/02/2025
									A0908	Incorrect, Inadequate or Imprecise Result or Readings	03/02/2025	03/02/2025
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	03/02/2025	03/02/2025
									A0711	Interrogation Problem	03/02/2025	03/02/2025
									A0908	Incorrect, Inadequate or Imprecise Result or Readings	03/02/2025	03/02/2025
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	03/02/2025	03/02/2025
									A0711	Interrogation Problem	03/02/2025	03/02/2025
									A0908	Incorrect, Inadequate or Imprecise Result or Readings	03/02/2025	03/02/2025
16585	2025	All other reportable incidents	47265	Dual-chamber implantable pacemaker, rate-responsive	CT1685	Cardiac pacemakers	E0112	Dizziness	A0712	Pacing Problem	29/01/2025	29/01/2025
									A0908	Incorrect, Inadequate or Imprecise Result or Readings	29/01/2025	29/01/2025
							F08	Hospitalization or Prolonged Hospitalization	A0712	Pacing Problem	29/01/2025	29/01/2025
									A0908	Incorrect, Inadequate or Imprecise Result or Readings	29/01/2025	29/01/2025
							F1001	Absence of Treatment	A0712	Pacing Problem	29/01/2025	29/01/2025
									A0908	Incorrect, Inadequate or Imprecise Result or Readings	29/01/2025	29/01/2025
							F1905	Device Revision or Replacement	A0712	Pacing Problem	29/01/2025	29/01/2025
									A0908	Incorrect, Inadequate or Imprecise Result or Readings	29/01/2025	29/01/2025
					CT257	Implantable cardiac pacemakers	E0112	Dizziness	A0712	Pacing Problem	29/01/2025	29/01/2025
									A0908	Incorrect, Inadequate or Imprecise Result or Readings	29/01/2025	29/01/2025
							F08	Hospitalization or Prolonged Hospitalization	A0712	Pacing Problem	29/01/2025	29/01/2025
									A0908	Incorrect, Inadequate or Imprecise Result or Readings	29/01/2025	29/01/2025
							F1001	Absence of Treatment	A0712	Pacing Problem	29/01/2025	29/01/2025
									A0908	Incorrect, Inadequate or Imprecise Result or Readings	29/01/2025	29/01/2025
							F1905	Device Revision or Replacement	A0712	Pacing Problem	29/01/2025	29/01/2025
									A0908	Incorrect, Inadequate or Imprecise Result or Readings	29/01/2025	29/01/2025
					CT2866	Implantable cardiac electrical stimulators	E0112	Dizziness	A0712	Pacing Problem	29/01/2025	29/01/2025
									A0908	Incorrect, Inadequate or Imprecise Result or Readings	29/01/2025	29/01/2025
							F08	Hospitalization or Prolonged Hospitalization	A0712	Pacing Problem	29/01/2025	29/01/2025
									A0908	Incorrect, Inadequate or Imprecise Result or Readings	29/01/2025	29/01/2025
							F1001	Absence of Treatment	A0712	Pacing Problem	29/01/2025	29/01/2025

									A0908	Incorrect, Inadequate or Imprecise Result or Readings	29/01/2025	29/01/2025
							F1905	Device Revision or Replacement	A0712	Pacing Problem	29/01/2025	29/01/2025
									A0908	Incorrect, Inadequate or Imprecise Result or Readings	29/01/2025	29/01/2025
16586	2025	All other reportable incidents	47270	Cardiac resynchronization therapy implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A072201	High Impedance		
							F26	No Health Consequences or Impact	A072201	High Impedance		
16587	2025	All other reportable incidents	35852	Single-chamber implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	07/05/2025	07/05/2025
									A0908	Incorrect, Inadequate or Imprecise Result or Readings	07/05/2025	07/05/2025
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	07/05/2025	07/05/2025
									A0908	Incorrect, Inadequate or Imprecise Result or Readings	07/05/2025	07/05/2025
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	07/05/2025	07/05/2025
									A0908	Incorrect, Inadequate or Imprecise Result or Readings	07/05/2025	07/05/2025
16588	2025	All other reportable incidents	17882	Physiologic-monitoring defibrillation system	CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A0719	Unexpected Shutdown		
							F26	No Health Consequences or Impact	A0719	Unexpected Shutdown		
16589	2025	All other reportable incidents	17882	Physiologic-monitoring defibrillation system	CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070603	Failure to Charge		
							F27	Problem Identified During Non-Clinical Procedure	A070603	Failure to Charge		
16590	2025	All other reportable incidents	35852	Single-chamber implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E020204	Malaise	A070909	Oversensing	22/08/2024	22/08/2024
									A070910	Undersensing	22/08/2024	22/08/2024
									A071302	Inappropriate/Inadequate Shock/Stimulation	22/08/2024	22/08/2024
									A090801	Signal Artifact/Noise	22/08/2024	22/08/2024
							E210401	Shock from Patient Lead(s)	A070909	Oversensing	22/08/2024	22/08/2024
									A070910	Undersensing	22/08/2024	22/08/2024
									A071302	Inappropriate/Inadequate Shock/Stimulation	22/08/2024	22/08/2024
									A090801	Signal Artifact/Noise	22/08/2024	22/08/2024
							F08	Hospitalization or Prolonged Hospitalization	A070909	Oversensing	22/08/2024	22/08/2024
									A070910	Undersensing	22/08/2024	22/08/2024
									A071302	Inappropriate/Inadequate Shock/Stimulation	22/08/2024	22/08/2024
									A090801	Signal Artifact/Noise	22/08/2024	22/08/2024
							F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A070909	Oversensing	22/08/2024	22/08/2024
									A070910	Undersensing	22/08/2024	22/08/2024
									A071302	Inappropriate/Inadequate Shock/Stimulation	22/08/2024	22/08/2024
									A090801	Signal Artifact/Noise	22/08/2024	22/08/2024
							F1905	Device Revision or Replacement	A070909	Oversensing	22/08/2024	22/08/2024
									A070910	Undersensing	22/08/2024	22/08/2024
									A071302	Inappropriate/Inadequate Shock/Stimulation	22/08/2024	22/08/2024
									A090801	Signal Artifact/Noise	22/08/2024	22/08/2024
							F23	Unexpected Medical Intervention	A070909	Oversensing	22/08/2024	22/08/2024
									A070910	Undersensing	22/08/2024	22/08/2024
									A071302	Inappropriate/Inadequate Shock/Stimulation	22/08/2024	22/08/2024
									A090801	Signal Artifact/Noise	22/08/2024	22/08/2024
16591	2025	All other reportable incidents	47267	Single-chamber implantable pacemaker, rate-responsive	CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem	10/01/2025	10/01/2025
									A070504	Premature Discharge of Battery	10/01/2025	10/01/2025
							F08	Hospitalization or Prolonged Hospitalization	A0705	Battery Problem	10/01/2025	10/01/2025
									A070504	Premature Discharge of Battery	10/01/2025	10/01/2025
							F1905	Device Revision or Replacement	A0705	Battery Problem	10/01/2025	10/01/2025
									A070504	Premature Discharge of Battery	10/01/2025	10/01/2025
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem	10/01/2025	10/01/2025
									A070504	Premature Discharge of Battery	10/01/2025	10/01/2025
							F08	Hospitalization or Prolonged Hospitalization	A0705	Battery Problem	10/01/2025	10/01/2025
									A070504	Premature Discharge of Battery	10/01/2025	10/01/2025
							F1905	Device Revision or Replacement	A0705	Battery Problem	10/01/2025	10/01/2025
									A070504	Premature Discharge of Battery	10/01/2025	10/01/2025
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem	10/01/2025	10/01/2025
									A070504	Premature Discharge of Battery	10/01/2025	10/01/2025
							F08	Hospitalization or Prolonged Hospitalization	A0705	Battery Problem	10/01/2025	10/01/2025
									A070504	Premature Discharge of Battery	10/01/2025	10/01/2025
							F1905	Device Revision or Replacement	A0705	Battery Problem	10/01/2025	10/01/2025
									A070504	Premature Discharge of Battery	10/01/2025	10/01/2025
16592	2025	All other reportable incidents			CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem	02/02/2025	02/02/2025
									A070504	Premature Discharge of Battery	02/02/2025	02/02/2025
							F08	Hospitalization or Prolonged Hospitalization	A0705	Battery Problem	02/02/2025	02/02/2025
									A070504	Premature Discharge of Battery	02/02/2025	02/02/2025
							F1905	Device Revision or Replacement	A0705	Battery Problem	02/02/2025	02/02/2025
									A070504	Premature Discharge of Battery	02/02/2025	02/02/2025
16593	2025	All other reportable incidents	47265	Dual-chamber implantable pacemaker, rate-responsive	CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem	19/03/2025	19/03/2025
									A070504	Premature Discharge of Battery	19/03/2025	19/03/2025
							F08	Hospitalization or Prolonged Hospitalization	A0705	Battery Problem	19/03/2025	19/03/2025
									A070504	Premature Discharge of Battery	19/03/2025	19/03/2025
							F1905	Device Revision or Replacement	A0705	Battery Problem	19/03/2025	19/03/2025
									A070504	Premature Discharge of Battery	19/03/2025	19/03/2025
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem	19/03/2025	19/03/2025
									A070504	Premature Discharge of Battery	19/03/2025	19/03/2025
							F08	Hospitalization or Prolonged Hospitalization	A0705	Battery Problem	19/03/2025	19/03/2025
									A070504	Premature Discharge of Battery	19/03/2025	19/03/2025
							F1905	Device Revision or Replacement	A0705	Battery Problem	19/03/2025	19/03/2025
									A070504	Premature Discharge of Battery	19/03/2025	19/03/2025
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem	19/03/2025	19/03/2025
									A070504	Premature Discharge of Battery	19/03/2025	19/03/2025
							F08	Hospitalization or Prolonged Hospitalization	A0705	Battery Problem	19/03/2025	19/03/2025
									A070504	Premature Discharge of Battery	19/03/2025	19/03/2025
							F1905	Device Revision or Replacement	A0705	Battery Problem	19/03/2025	19/03/2025
									A070504	Premature Discharge of Battery	19/03/2025	19/03/2025
16594	2025	All other reportable incidents	47265	Dual-chamber implantable pacemaker, rate-responsive	CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A0908	Incorrect, Inadequate or Imprecise Result or Readings		

							F26	No Health Consequences or Impact	A0908	Incorrect, Inadequate or Imprecise Result or Readings			
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A0908	Incorrect, Inadequate or Imprecise Result or Readings			
							F26	No Health Consequences or Impact	A0908	Incorrect, Inadequate or Imprecise Result or Readings			
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A0908	Incorrect, Inadequate or Imprecise Result or Readings			
							F26	No Health Consequences or Impact	A0908	Incorrect, Inadequate or Imprecise Result or Readings			
16595	2025	All other reportable incidents	17882	Physiologic-monitoring defibrillation system	CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070803	Failure to Power Up			
									A150105	Key or Button Unresponsive/not Working			
							F26	No Health Consequences or Impact	A070803	Failure to Power Up			
									A150105	Key or Button Unresponsive/not Working			
16596	2025	All other reportable incidents	17882	Physiologic-monitoring defibrillation system	CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A1601	Device Alarm System			
							F27	Problem Identified During Non-Clinical Procedure	A1601	Device Alarm System			
16597	2025	All other reportable incidents	37265	Dual-chamber implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery			
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery			
16598	2025	All other reportable incidents	17882	Physiologic-monitoring defibrillation system	CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A090811	Unable to Obtain Readings			
							F27	Problem Identified During Non-Clinical Procedure	A090811	Unable to Obtain Readings			
16599	2025	All other reportable incidents	47205	Cardiac pulse generator programmer	CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A0902	Display or Visual Feedback Problem			
							F26	No Health Consequences or Impact	A0902	Display or Visual Feedback Problem			
16600	2025	All other reportable incidents	47265	Dual-chamber implantable pacemaker, rate-responsive	CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A090801	Signal Artifact/Noise	03/02/2025	03/02/2025	
									A12	Connection Problem	03/02/2025	03/02/2025	
							F08	Hospitalization or Prolonged Hospitalization	A090801	Signal Artifact/Noise	03/02/2025	03/02/2025	
									A12	Connection Problem	03/02/2025	03/02/2025	
							F1908	Prolonged Surgery	A090801	Signal Artifact/Noise	03/02/2025	03/02/2025	
									A12	Connection Problem	03/02/2025	03/02/2025	
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A090801	Signal Artifact/Noise	03/02/2025	03/02/2025	
									A12	Connection Problem	03/02/2025	03/02/2025	
							F08	Hospitalization or Prolonged Hospitalization	A090801	Signal Artifact/Noise	03/02/2025	03/02/2025	
									A12	Connection Problem	03/02/2025	03/02/2025	
							F1908	Prolonged Surgery	A090801	Signal Artifact/Noise	03/02/2025	03/02/2025	
									A12	Connection Problem	03/02/2025	03/02/2025	
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A090801	Signal Artifact/Noise	03/02/2025	03/02/2025	
									A12	Connection Problem	03/02/2025	03/02/2025	
							F08	Hospitalization or Prolonged Hospitalization	A090801	Signal Artifact/Noise	03/02/2025	03/02/2025	
									A12	Connection Problem	03/02/2025	03/02/2025	
							F1908	Prolonged Surgery	A090801	Signal Artifact/Noise	03/02/2025	03/02/2025	
									A12	Connection Problem	03/02/2025	03/02/2025	
16601	2025	All other reportable incidents	47270	Cardiac resynchronization therapy implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A070909	Oversensing			
									A090801	Signal Artifact/Noise			
									A1301	Failure to Read Input Signal			
							F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A070909	Oversensing			
									A090801	Signal Artifact/Noise			
									A1301	Failure to Read Input Signal			
16602	2025	All other reportable incidents	47265	Dual-chamber implantable pacemaker, rate-responsive	CT1685	Cardiac pacemakers	E2103	Device Overstimulation of Tissue	A27	Appropriate Device Problem Term/Code not Available			
							F12	Serious Injury/ Illness/ Impairment	A27	Appropriate Device Problem Term/Code not Available			
							F2307	Reprogramming of Device	A27	Appropriate Device Problem Term/Code not Available			
					CT257	Implantable cardiac pacemakers	E2103	Device Overstimulation of Tissue	A27	Appropriate Device Problem Term/Code not Available			
							F12	Serious Injury/ Illness/ Impairment	A27	Appropriate Device Problem Term/Code not Available			
							F2307	Reprogramming of Device	A27	Appropriate Device Problem Term/Code not Available			
					CT2866	Implantable cardiac electrical stimulators	E2103	Device Overstimulation of Tissue	A27	Appropriate Device Problem Term/Code not Available			
							F12	Serious Injury/ Illness/ Impairment	A27	Appropriate Device Problem Term/Code not Available			
							F2307	Reprogramming of Device	A27	Appropriate Device Problem Term/Code not Available			
16603	2025	All other reportable incidents	37265	Dual-chamber implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	06/03/2025	06/03/2025	
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	06/03/2025	06/03/2025	
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	06/03/2025	06/03/2025	
16604	2025	All other reportable incidents	47270	Cardiac resynchronization therapy implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A070909	Oversensing	19/11/2024	19/11/2024	
									A090801	Signal Artifact/Noise	19/11/2024	19/11/2024	
							F08	Hospitalization or Prolonged Hospitalization	A070909	Oversensing	19/11/2024	19/11/2024	
									A090801	Signal Artifact/Noise	19/11/2024	19/11/2024	
							F1905	Device Revision or Replacement	A070909	Oversensing	19/11/2024	19/11/2024	
									A090801	Signal Artifact/Noise	19/11/2024	19/11/2024	
16605	2025	All other reportable incidents	47265	Dual-chamber implantable pacemaker, rate-responsive	CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem	31/01/2025	31/01/2025	
									A070504	Premature Discharge of Battery	31/01/2025	31/01/2025	
							F08	Hospitalization or Prolonged Hospitalization	A0705	Battery Problem	31/01/2025	31/01/2025	
									A070504	Premature Discharge of Battery	31/01/2025	31/01/2025	
							F1905	Device Revision or Replacement	A0705	Battery Problem	31/01/2025	31/01/2025	
									A070504	Premature Discharge of Battery	31/01/2025	31/01/2025	
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem	31/01/2025	31/01/2025	
									A070504	Premature Discharge of Battery	31/01/2025	31/01/2025	
							F08	Hospitalization or Prolonged Hospitalization	A0705	Battery Problem	31/01/2025	31/01/2025	
									A070504	Premature Discharge of Battery	31/01/2025	31/01/2025	
							F1905	Device Revision or Replacement	A0705	Battery Problem	31/01/2025	31/01/2025	
									A070504	Premature Discharge of Battery	31/01/2025	31/01/2025	
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem	31/01/2025	31/01/2025	
									A070504	Premature Discharge of Battery	31/01/2025	31/01/2025	
							F08	Hospitalization or Prolonged Hospitalization	A0705	Battery Problem	31/01/2025	31/01/2025	
									A070504	Premature Discharge of Battery	31/01/2025	31/01/2025	
							F1905	Device Revision or Replacement	A0705	Battery Problem	31/01/2025	31/01/2025	
									A070504	Premature Discharge of Battery	31/01/2025	31/01/2025	

16606	2025		60789	Intracardiac pacemaker	CT1685	Cardiac pacemakers	E0601	Arrhythmia	A24	No Apparent Device Problem		
							F02	Death	A24	No Apparent Device Problem		
					CT257	Implantable cardiac pacemakers	E0601	Arrhythmia	A24	No Apparent Device Problem		
							F02	Death	A24	No Apparent Device Problem		
					CT2818	Catheterization/Cannulation/Fluid path devices	E0601	Arrhythmia	A24	No Apparent Device Problem		
							F02	Death	A24	No Apparent Device Problem		
					CT2866	Implantable cardiac electrical stimulators	E0601	Arrhythmia	A24	No Apparent Device Problem		
							F02	Death	A24	No Apparent Device Problem		
16607	2025		60789	Intracardiac pacemaker	CT1685	Cardiac pacemakers	E0601	Arrhythmia	A24	No Apparent Device Problem		
							F02	Death	A24	No Apparent Device Problem		
					CT257	Implantable cardiac pacemakers	E0601	Arrhythmia	A24	No Apparent Device Problem		
							F02	Death	A24	No Apparent Device Problem		
					CT2818	Catheterization/Cannulation/Fluid path devices	E0601	Arrhythmia	A24	No Apparent Device Problem		
							F02	Death	A24	No Apparent Device Problem		
					CT2866	Implantable cardiac electrical stimulators	E0601	Arrhythmia	A24	No Apparent Device Problem		
							F02	Death	A24	No Apparent Device Problem		
16608	2025	All other reportable incidents	47263	Cardiac resynchronization therapy implantable pacemaker	CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
16609	2025	All other reportable incidents	47265	Dual-chamber implantable pacemaker, rate-responsive	CT1685	Cardiac pacemakers	E0601	Arrhythmia	A071102	Failure to Interrogate	25/02/2025	25/02/2025
									A0712	Pacing Problem	25/02/2025	25/02/2025
							F1901	Additional Surgery	A071102	Failure to Interrogate	25/02/2025	25/02/2025
									A0712	Pacing Problem	25/02/2025	25/02/2025
					CT257	Implantable cardiac pacemakers	E0601	Arrhythmia	A071102	Failure to Interrogate	25/02/2025	25/02/2025
									A0712	Pacing Problem	25/02/2025	25/02/2025
							F1901	Additional Surgery	A071102	Failure to Interrogate	25/02/2025	25/02/2025
									A0712	Pacing Problem	25/02/2025	25/02/2025
					CT2866	Implantable cardiac electrical stimulators	E0601	Arrhythmia	A071102	Failure to Interrogate	25/02/2025	25/02/2025
									A0712	Pacing Problem	25/02/2025	25/02/2025
							F1901	Additional Surgery	A071102	Failure to Interrogate	25/02/2025	25/02/2025
									A0712	Pacing Problem	25/02/2025	25/02/2025
16610	2025	All other reportable incidents			CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070909	Oversensing		
							F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A070909	Oversensing		
16611	2025	All other reportable incidents	37265	Dual-chamber implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A072201	High Impedance		
							F26	No Health Consequences or Impact	A072201	High Impedance		
16612	2025	All other reportable incidents	47270	Cardiac resynchronization therapy implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A010402	Migration		
									A070909	Oversensing		
									A070910	Undersensing		
									A1301	Failure to Read Input Signal		
							F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A010402	Migration		
									A070909	Oversensing		
									A070910	Undersensing		
									A1301	Failure to Read Input Signal		
							F2203	Imaging Required	A010402	Migration		
									A070909	Oversensing		
									A070910	Undersensing		
									A1301	Failure to Read Input Signal		
							F23	Unexpected Medical Intervention	A010402	Migration		
									A070909	Oversensing		
									A070910	Undersensing		
									A1301	Failure to Read Input Signal		
16613	2025	All other reportable incidents	47205	Cardiac pulse generator programmer	CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A0902	Display or Visual Feedback Problem		
							F27	Problem Identified During Non-Clinical Procedure	A0902	Display or Visual Feedback Problem		
16614	2025		47265	Dual-chamber implantable pacemaker, rate-responsive	CT1685	Cardiac pacemakers	E020203	Distress	A070504	Premature Discharge of Battery		
							F1901	Additional Surgery	A070504	Premature Discharge of Battery		
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery		
					CT257	Implantable cardiac pacemakers	E020203	Distress	A070504	Premature Discharge of Battery		
							F1901	Additional Surgery	A070504	Premature Discharge of Battery		
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery		
					CT2866	Implantable cardiac electrical stimulators	E020203	Distress	A070504	Premature Discharge of Battery		
							F1901	Additional Surgery	A070504	Premature Discharge of Battery		
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery		
16615	2025	All other reportable incidents	47265	Dual-chamber implantable pacemaker, rate-responsive	CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	06/01/2025	06/01/2025
									A0908	Incorrect, Inadequate or Imprecise Result or Readings	06/01/2025	06/01/2025
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	06/01/2025	06/01/2025
									A0908	Incorrect, Inadequate or Imprecise Result or Readings	06/01/2025	06/01/2025
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	06/01/2025	06/01/2025
									A0908	Incorrect, Inadequate or Imprecise Result or Readings	06/01/2025	06/01/2025
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	06/01/2025	06/01/2025
									A0908	Incorrect, Inadequate or Imprecise Result or Readings	06/01/2025	06/01/2025
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	06/01/2025	06/01/2025
									A0908	Incorrect, Inadequate or Imprecise Result or Readings	06/01/2025	06/01/2025
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	06/01/2025	06/01/2025
									A0908	Incorrect, Inadequate or Imprecise Result or Readings	06/01/2025	06/01/2025
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	06/01/2025	06/01/2025
									A0908	Incorrect, Inadequate or Imprecise Result or Readings	06/01/2025	06/01/2025
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	06/01/2025	06/01/2025
									A0908	Incorrect, Inadequate or Imprecise Result or Readings	06/01/2025	06/01/2025
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	06/01/2025	06/01/2025

									A0908	Incorrect, Inadequate or Imprecise Result or Readings	06/01/2025	06/01/2025
16616	2025	All other reportable incidents	47263	Cardiac resynchronization therapy implantable pacemaker	CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A0908	Incorrect, Inadequate or Imprecise Result or Readings	24/03/2025	24/03/2025
							F08	Hospitalization or Prolonged Hospitalization	A0908	Incorrect, Inadequate or Imprecise Result or Readings	24/03/2025	24/03/2025
							F1905	Device Revision or Replacement	A0908	Incorrect, Inadequate or Imprecise Result or Readings	24/03/2025	24/03/2025
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A0908	Incorrect, Inadequate or Imprecise Result or Readings	24/03/2025	24/03/2025
							F08	Hospitalization or Prolonged Hospitalization	A0908	Incorrect, Inadequate or Imprecise Result or Readings	24/03/2025	24/03/2025
							F1905	Device Revision or Replacement	A0908	Incorrect, Inadequate or Imprecise Result or Readings	24/03/2025	24/03/2025
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A0908	Incorrect, Inadequate or Imprecise Result or Readings	24/03/2025	24/03/2025
							F08	Hospitalization or Prolonged Hospitalization	A0908	Incorrect, Inadequate or Imprecise Result or Readings	24/03/2025	24/03/2025
							F1905	Device Revision or Replacement	A0908	Incorrect, Inadequate or Imprecise Result or Readings	24/03/2025	24/03/2025
16617	2025	All other reportable incidents	37265	Dual-chamber implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E210401	Shock from Patient Lead(s)	A071302	Inappropriate/Inadequate Shock/Stimulation		
							F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A071302	Inappropriate/Inadequate Shock/Stimulation		
16618	2025	All other reportable incidents	17882	Physiologic-monitoring defibrillation system	CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A090201	Device Displays Incorrect Message		
							F26	No Health Consequences or Impact	A090201	Device Displays Incorrect Message		
16619	2025	All other reportable incidents	47265	Dual-chamber implantable pacemaker, rate-responsive	CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A072202	Low Impedance		
							F26	No Health Consequences or Impact	A072202	Low Impedance		
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A072202	Low Impedance		
							F26	No Health Consequences or Impact	A072202	Low Impedance		
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A072202	Low Impedance		
							F26	No Health Consequences or Impact	A072202	Low Impedance		
16620	2025	All other reportable incidents	47205	Cardiac pulse generator programmer	CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A0902	Display or Visual Feedback Problem		
							F27	Problem Identified During Non-Clinical Procedure	A0902	Display or Visual Feedback Problem		
16621	2025		60956	Pacing/defibrillation lead tunneller	CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem		
									A160105	False Alarm		
							F26	No Health Consequences or Impact	A0705	Battery Problem		
									A160105	False Alarm		
16622	2025	All other reportable incidents	47265	Dual-chamber implantable pacemaker, rate-responsive	CT1685	Cardiac pacemakers	E0601	Arrhythmia	A070101	Failure to Capture		
							F26	No Health Consequences or Impact	A070101	Failure to Capture		
					CT257	Implantable cardiac pacemakers	E0601	Arrhythmia	A070101	Failure to Capture		
							F26	No Health Consequences or Impact	A070101	Failure to Capture		
					CT2866	Implantable cardiac electrical stimulators	E0601	Arrhythmia	A070101	Failure to Capture		
							F26	No Health Consequences or Impact	A070101	Failure to Capture		
16623	2025	All other reportable incidents	47263	Cardiac resynchronization therapy implantable pacemaker	CT1685	Cardiac pacemakers	E0112	Dizziness	A070909	Oversensing	31/01/2025	31/01/2025
									A0908	Incorrect, Inadequate or Imprecise Result or Readings	31/01/2025	31/01/2025
							E060104	Bradycardia	A070909	Oversensing	31/01/2025	31/01/2025
									A0908	Incorrect, Inadequate or Imprecise Result or Readings	31/01/2025	31/01/2025
							F08	Hospitalization or Prolonged Hospitalization	A070909	Oversensing	31/01/2025	31/01/2025
									A0908	Incorrect, Inadequate or Imprecise Result or Readings	31/01/2025	31/01/2025
							F1905	Device Revision or Replacement	A070909	Oversensing	31/01/2025	31/01/2025
									A0908	Incorrect, Inadequate or Imprecise Result or Readings	31/01/2025	31/01/2025
					CT257	Implantable cardiac pacemakers	E0112	Dizziness	A070909	Oversensing	31/01/2025	31/01/2025
									A0908	Incorrect, Inadequate or Imprecise Result or Readings	31/01/2025	31/01/2025
							E060104	Bradycardia	A070909	Oversensing	31/01/2025	31/01/2025
									A0908	Incorrect, Inadequate or Imprecise Result or Readings	31/01/2025	31/01/2025
							F08	Hospitalization or Prolonged Hospitalization	A070909	Oversensing	31/01/2025	31/01/2025
									A0908	Incorrect, Inadequate or Imprecise Result or Readings	31/01/2025	31/01/2025
							F1905	Device Revision or Replacement	A070909	Oversensing	31/01/2025	31/01/2025
									A0908	Incorrect, Inadequate or Imprecise Result or Readings	31/01/2025	31/01/2025
					CT2866	Implantable cardiac electrical stimulators	E0112	Dizziness	A070909	Oversensing	31/01/2025	31/01/2025
									A0908	Incorrect, Inadequate or Imprecise Result or Readings	31/01/2025	31/01/2025
							E060104	Bradycardia	A070909	Oversensing	31/01/2025	31/01/2025
									A0908	Incorrect, Inadequate or Imprecise Result or Readings	31/01/2025	31/01/2025
							F08	Hospitalization or Prolonged Hospitalization	A070909	Oversensing	31/01/2025	31/01/2025
									A0908	Incorrect, Inadequate or Imprecise Result or Readings	31/01/2025	31/01/2025
							F1905	Device Revision or Replacement	A070909	Oversensing	31/01/2025	31/01/2025
									A0908	Incorrect, Inadequate or Imprecise Result or Readings	31/01/2025	31/01/2025
16624	2025	All other reportable incidents	47270	Cardiac resynchronization therapy implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E060109	Tachycardia	A0709	Device Sensing Problem		
									A0712	Pacing Problem		
							F11	Non-Serious Injury/ Illness/ Impairment	A0709	Device Sensing Problem		
									A0712	Pacing Problem		
							F12	Serious Injury/ Illness/ Impairment	A0709	Device Sensing Problem		
									A0712	Pacing Problem		
							F2307	Reprogramming of Device	A0709	Device Sensing Problem		
									A0712	Pacing Problem		
16625	2025	All other reportable incidents	47270	Cardiac resynchronization therapy implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A070909	Oversensing		
							F05	Delay to Treatment/ Therapy	A070909	Oversensing		
16626	2025	All other reportable incidents	47270	Cardiac resynchronization therapy implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	05/06/2025	05/06/2025
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	05/06/2025	05/06/2025
							F1903	Device Explantation	A070504	Premature Discharge of Battery	05/06/2025	05/06/2025
16627	2025	All other reportable incidents	63188	Left-ventricular wireless pacing system sensor/transmitter	CT145	Cardiovascular devices			A0708	Power Problem		
					CT1577	Wireless monitoring system transmitters			A0708	Power Problem		

					CT1684	Cardiac pacemakers and associated devices			A0708	Power Problem		
					CT2727	Active implant devices/ancillaries			A0708	Power Problem		
					CT2750	Implantable cardiac electrical stimulators and associated devices			A0708	Power Problem		
					CT2885	Physiological monitoring/measuring system transmitters/receivers/transceivers			A0708	Power Problem		
16628	2025	All other reportable incidents	47270	Cardiac resynchronization therapy implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	07/11/2024	07/11/2024
									A0908	Incorrect, Inadequate or Imprecise Result or Readings	07/11/2024	07/11/2024
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	07/11/2024	07/11/2024
									A0908	Incorrect, Inadequate or Imprecise Result or Readings	07/11/2024	07/11/2024
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	07/11/2024	07/11/2024
									A0908	Incorrect, Inadequate or Imprecise Result or Readings	07/11/2024	07/11/2024
16629	2025	All other reportable incidents	47205	Cardiac pulse generator programmer	CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A0902	Display or Visual Feedback Problem		
							F27	Problem Identified During Non-Clinical Procedure	A0902	Display or Visual Feedback Problem		
16630	2025	All other reportable incidents	47263	Cardiac resynchronization therapy implantable pacemaker	CT1685	Cardiac pacemakers	E2103	Device Overstimulation of Tissue	A0908	Incorrect, Inadequate or Imprecise Result or Readings	06/02/2025	06/02/2025
							F08	Hospitalization or Prolonged Hospitalization	A0908	Incorrect, Inadequate or Imprecise Result or Readings	06/02/2025	06/02/2025
							F1905	Device Revision or Replacement	A0908	Incorrect, Inadequate or Imprecise Result or Readings	06/02/2025	06/02/2025
					CT257	Implantable cardiac pacemakers	E2103	Device Overstimulation of Tissue	A0908	Incorrect, Inadequate or Imprecise Result or Readings	06/02/2025	06/02/2025
							F08	Hospitalization or Prolonged Hospitalization	A0908	Incorrect, Inadequate or Imprecise Result or Readings	06/02/2025	06/02/2025
							F1905	Device Revision or Replacement	A0908	Incorrect, Inadequate or Imprecise Result or Readings	06/02/2025	06/02/2025
					CT2866	Implantable cardiac electrical stimulators	E2103	Device Overstimulation of Tissue	A0908	Incorrect, Inadequate or Imprecise Result or Readings	06/02/2025	06/02/2025
							F08	Hospitalization or Prolonged Hospitalization	A0908	Incorrect, Inadequate or Imprecise Result or Readings	06/02/2025	06/02/2025
							F1905	Device Revision or Replacement	A0908	Incorrect, Inadequate or Imprecise Result or Readings	06/02/2025	06/02/2025
16631	2025	All other reportable incidents	47265	Dual-chamber implantable pacemaker, rate-responsive	CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A05	Mechanical Problem	10/02/2025	10/02/2025
							F08	Hospitalization or Prolonged Hospitalization	A05	Mechanical Problem	10/02/2025	10/02/2025
							F12	Serious Injury/ Illness/ Impairment	A05	Mechanical Problem	10/02/2025	10/02/2025
							F1905	Device Revision or Replacement	A05	Mechanical Problem	10/02/2025	10/02/2025
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A05	Mechanical Problem	10/02/2025	10/02/2025
							F08	Hospitalization or Prolonged Hospitalization	A05	Mechanical Problem	10/02/2025	10/02/2025
							F12	Serious Injury/ Illness/ Impairment	A05	Mechanical Problem	10/02/2025	10/02/2025
							F1905	Device Revision or Replacement	A05	Mechanical Problem	10/02/2025	10/02/2025
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A05	Mechanical Problem	10/02/2025	10/02/2025
							F08	Hospitalization or Prolonged Hospitalization	A05	Mechanical Problem	10/02/2025	10/02/2025
							F12	Serious Injury/ Illness/ Impairment	A05	Mechanical Problem	10/02/2025	10/02/2025
							F1905	Device Revision or Replacement	A05	Mechanical Problem	10/02/2025	10/02/2025
16632	2025		47265	Dual-chamber implantable pacemaker, rate-responsive	CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A0203	Defective Device		
							F26	No Health Consequences or Impact	A0203	Defective Device		
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A0203	Defective Device		
							F26	No Health Consequences or Impact	A0203	Defective Device		
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A0203	Defective Device		
							F26	No Health Consequences or Impact	A0203	Defective Device		
16633	2025		37265	Dual-chamber implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem		
							F1905	Device Revision or Replacement	A0705	Battery Problem		
16634	2025		60956	Pacing/defibrillation lead tunneller	CT1684	Cardiac pacemakers and associated devices	E210401	Shock from Patient Lead(s)	A071302	Inappropriate/Inadequate Shock/Stimulation		
									A090801	Signal Artifact/Noise		
							F1903	Device Explantation	A071302	Inappropriate/Inadequate Shock/Stimulation		
									A090801	Signal Artifact/Noise		
16635	2025		60789	Intracardiac pacemaker	CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A0201	Product Quality Problem		
									A140801	Failure to Deliver		
							F1905	Device Revision or Replacement	A0201	Product Quality Problem		
									A140801	Failure to Deliver		
							F26	No Health Consequences or Impact	A0201	Product Quality Problem		
									A140801	Failure to Deliver		
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A0201	Product Quality Problem		
									A140801	Failure to Deliver		
							F1905	Device Revision or Replacement	A0201	Product Quality Problem		
									A140801	Failure to Deliver		
							F26	No Health Consequences or Impact	A0201	Product Quality Problem		
									A140801	Failure to Deliver		
					CT2818	Catheterization/Cannulation/Fluid path devices	E2403	No Clinical Signs, Symptoms or Conditions	A0201	Product Quality Problem		
									A140801	Failure to Deliver		
							F1905	Device Revision or Replacement	A0201	Product Quality Problem		
									A140801	Failure to Deliver		
							F26	No Health Consequences or Impact	A0201	Product Quality Problem		
									A140801	Failure to Deliver		
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A0201	Product Quality Problem		
									A140801	Failure to Deliver		
							F1905	Device Revision or Replacement	A0201	Product Quality Problem		
									A140801	Failure to Deliver		
							F26	No Health Consequences or Impact	A0201	Product Quality Problem		
									A140801	Failure to Deliver		
16636	2025		37265	Dual-chamber implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem		
							F1905	Device Revision or Replacement	A0705	Battery Problem		
16637	2025		47265	Dual-chamber implantable pacemaker, rate-responsive	CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A0203	Defective Device		
									A070102	High Capture Threshold		
							F1905	Device Revision or Replacement	A0203	Defective Device		
									A070102	High Capture Threshold		
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A0203	Defective Device		
									A070102	High Capture Threshold		

							F1905	Device Revision or Replacement	A0203	Defective Device		
									A070102	High Capture Threshold		
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A0203	Defective Device		
									A070102	High Capture Threshold		
							F1905	Device Revision or Replacement	A0203	Defective Device		
									A070102	High Capture Threshold		
16638	2025	All other reportable incidents	47265	Dual-chamber implantable pacemaker, rate-responsive	CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	26/02/2025	26/02/2025
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	26/02/2025	26/02/2025
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	26/02/2025	26/02/2025
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	26/02/2025	26/02/2025
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	26/02/2025	26/02/2025
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	26/02/2025	26/02/2025
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	26/02/2025	26/02/2025
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	26/02/2025	26/02/2025
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	26/02/2025	26/02/2025
16639	2025	All other reportable incidents	47265	Dual-chamber implantable pacemaker, rate-responsive	CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A0908	Incorrect, Inadequate or Imprecise Result or Readings	11/02/2025	11/02/2025
							F08	Hospitalization or Prolonged Hospitalization	A0908	Incorrect, Inadequate or Imprecise Result or Readings	11/02/2025	11/02/2025
							F1905	Device Revision or Replacement	A0908	Incorrect, Inadequate or Imprecise Result or Readings	11/02/2025	11/02/2025
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A0908	Incorrect, Inadequate or Imprecise Result or Readings	11/02/2025	11/02/2025
							F08	Hospitalization or Prolonged Hospitalization	A0908	Incorrect, Inadequate or Imprecise Result or Readings	11/02/2025	11/02/2025
							F1905	Device Revision or Replacement	A0908	Incorrect, Inadequate or Imprecise Result or Readings	11/02/2025	11/02/2025
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A0908	Incorrect, Inadequate or Imprecise Result or Readings	11/02/2025	11/02/2025
							F08	Hospitalization or Prolonged Hospitalization	A0908	Incorrect, Inadequate or Imprecise Result or Readings	11/02/2025	11/02/2025
							F1905	Device Revision or Replacement	A0908	Incorrect, Inadequate or Imprecise Result or Readings	11/02/2025	11/02/2025
16640	2025	All other reportable incidents	47265	Dual-chamber implantable pacemaker, rate-responsive	CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
16641	2025	All other reportable incidents	47263	Cardiac resynchronization therapy implantable pacemaker	CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A0908	Incorrect, Inadequate or Imprecise Result or Readings		
							F26	No Health Consequences or Impact	A0908	Incorrect, Inadequate or Imprecise Result or Readings		
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A0908	Incorrect, Inadequate or Imprecise Result or Readings		
							F26	No Health Consequences or Impact	A0908	Incorrect, Inadequate or Imprecise Result or Readings		
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A0908	Incorrect, Inadequate or Imprecise Result or Readings		
							F26	No Health Consequences or Impact	A0908	Incorrect, Inadequate or Imprecise Result or Readings		
16642	2025	All other reportable incidents	47270	Cardiac resynchronization therapy implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	12/06/2025	12/06/2025
									A0908	Incorrect, Inadequate or Imprecise Result or Readings	12/06/2025	12/06/2025
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	12/06/2025	12/06/2025
									A0908	Incorrect, Inadequate or Imprecise Result or Readings	12/06/2025	12/06/2025
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	12/06/2025	12/06/2025
									A0908	Incorrect, Inadequate or Imprecise Result or Readings	12/06/2025	12/06/2025
16643	2025	All other reportable incidents	17882	Physiologic-monitoring defibrillation system	CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A12	Connection Problem		
							F26	No Health Consequences or Impact	A12	Connection Problem		
16644	2025	All other reportable incidents	36106	Implantable pacemaker analyser, intraoperative	CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A0902	Display or Visual Feedback Problem		
									A11	Computer Software Problem		
							F27	Problem Identified During Non-Clinical Procedure	A0902	Display or Visual Feedback Problem		
									A11	Computer Software Problem		
16645	2025	All other reportable incidents	36106	Implantable pacemaker analyser, intraoperative	CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A0902	Display or Visual Feedback Problem		
									A11	Computer Software Problem		
							F27	Problem Identified During Non-Clinical Procedure	A0902	Display or Visual Feedback Problem		
									A11	Computer Software Problem		
16646	2025	All other reportable incidents	47265	Dual-chamber implantable pacemaker, rate-responsive	CT1685	Cardiac pacemakers	E0601	Arrhythmia	A0712	Pacing Problem		
							F11	Non-Serious Injury/ Illness/ Impairment	A0712	Pacing Problem		
					CT257	Implantable cardiac pacemakers	E0601	Arrhythmia	A0712	Pacing Problem		
							F11	Non-Serious Injury/ Illness/ Impairment	A0712	Pacing Problem		
					CT2866	Implantable cardiac electrical stimulators	E0601	Arrhythmia	A0712	Pacing Problem		
							F11	Non-Serious Injury/ Illness/ Impairment	A0712	Pacing Problem		
16647	2025	All other reportable incidents	60789	Intracardiac pacemaker	CT1685	Cardiac pacemakers	E0619	Pericardial Effusion	A070101	Failure to Capture		
									A070102	High Capture Threshold		
									A070901	Decreased Sensitivity		
							E2321	Low Blood Pressure/ Hypotension	A070101	Failure to Capture		
									A070102	High Capture Threshold		
									A070901	Decreased Sensitivity		
							F12	Serious Injury/ Illness/ Impairment	A070101	Failure to Capture		
									A070102	High Capture Threshold		
									A070901	Decreased Sensitivity		
							F1905	Device Revision or Replacement	A070101	Failure to Capture		

									A070102	High Capture Threshold		
									A070901	Decreased Sensitivity		
							F2307	Reprogramming of Device	A070101	Failure to Capture		
									A070102	High Capture Threshold		
									A070901	Decreased Sensitivity		
					CT257	Implantable cardiac pacemakers	E0619	Pericardial Effusion	A070101	Failure to Capture		
									A070102	High Capture Threshold		
									A070901	Decreased Sensitivity		
							E2321	Low Blood Pressure/ Hypotension	A070101	Failure to Capture		
									A070102	High Capture Threshold		
									A070901	Decreased Sensitivity		
							F12	Serious Injury/ Illness/ Impairment	A070101	Failure to Capture		
									A070102	High Capture Threshold		
									A070901	Decreased Sensitivity		
							F1905	Device Revision or Replacement	A070101	Failure to Capture		
									A070102	High Capture Threshold		
									A070901	Decreased Sensitivity		
							F2307	Reprogramming of Device	A070101	Failure to Capture		
									A070102	High Capture Threshold		
									A070901	Decreased Sensitivity		
					CT2818	Catheterization/Cannulation/Fluid path devices	E0619	Pericardial Effusion	A070101	Failure to Capture		
									A070102	High Capture Threshold		
									A070901	Decreased Sensitivity		
							E2321	Low Blood Pressure/ Hypotension	A070101	Failure to Capture		
									A070102	High Capture Threshold		
									A070901	Decreased Sensitivity		
							F12	Serious Injury/ Illness/ Impairment	A070101	Failure to Capture		
									A070102	High Capture Threshold		
									A070901	Decreased Sensitivity		
							F1905	Device Revision or Replacement	A070101	Failure to Capture		
									A070102	High Capture Threshold		
									A070901	Decreased Sensitivity		
							F2307	Reprogramming of Device	A070101	Failure to Capture		
									A070102	High Capture Threshold		
									A070901	Decreased Sensitivity		
					CT2866	Implantable cardiac electrical stimulators	E0619	Pericardial Effusion	A070101	Failure to Capture		
									A070102	High Capture Threshold		
									A070901	Decreased Sensitivity		
							E2321	Low Blood Pressure/ Hypotension	A070101	Failure to Capture		
									A070102	High Capture Threshold		
									A070901	Decreased Sensitivity		
							F12	Serious Injury/ Illness/ Impairment	A070101	Failure to Capture		
									A070102	High Capture Threshold		
									A070901	Decreased Sensitivity		
							F1905	Device Revision or Replacement	A070101	Failure to Capture		
									A070102	High Capture Threshold		
									A070901	Decreased Sensitivity		
							F2307	Reprogramming of Device	A070101	Failure to Capture		
									A070102	High Capture Threshold		
									A070901	Decreased Sensitivity		
16648	2025	All other reportable incidents	60789	Intracardiac pacemaker	CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A051201	Device Dislodged or Dislocated		
							F08	Hospitalization or Prolonged Hospitalization	A051201	Device Dislodged or Dislocated		
							F1905	Device Revision or Replacement	A051201	Device Dislodged or Dislocated		
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A051201	Device Dislodged or Dislocated		
							F08	Hospitalization or Prolonged Hospitalization	A051201	Device Dislodged or Dislocated		
							F1905	Device Revision or Replacement	A051201	Device Dislodged or Dislocated		
					CT2818	Catheterization/Cannulation/Fluid path devices	E2403	No Clinical Signs, Symptoms or Conditions	A051201	Device Dislodged or Dislocated		
							F08	Hospitalization or Prolonged Hospitalization	A051201	Device Dislodged or Dislocated		
							F1905	Device Revision or Replacement	A051201	Device Dislodged or Dislocated		
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A051201	Device Dislodged or Dislocated		
							F08	Hospitalization or Prolonged Hospitalization	A051201	Device Dislodged or Dislocated		
							F1905	Device Revision or Replacement	A051201	Device Dislodged or Dislocated		
16649	2025	All other reportable incidents	47270	Cardiac resynchronization therapy implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E210401	Shock from Patient Lead(s)	A070909	Oversensing		
									A071302	Inappropriate/Inadequate Shock/Stimulation		
							F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A070909	Oversensing		
									A071302	Inappropriate/Inadequate Shock/Stimulation		
16650	2025	All other reportable incidents	47270	Cardiac resynchronization therapy implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	05/03/2025	05/03/2025
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	05/03/2025	05/03/2025
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	05/03/2025	05/03/2025
16651	2025	All other reportable incidents	47270	Cardiac resynchronization therapy implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	15/04/2025	15/04/2025
									A0908	Incorrect, Inadequate or Imprecise Result or Readings	15/04/2025	15/04/2025
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	15/04/2025	15/04/2025
									A0908	Incorrect, Inadequate or Imprecise Result or Readings	15/04/2025	15/04/2025
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	15/04/2025	15/04/2025
									A0908	Incorrect, Inadequate or Imprecise Result or Readings	15/04/2025	15/04/2025
16652	2025	All other reportable incidents	47205	Cardiac pulse generator programmer	CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A13	Communication or Transmission Problem		
							F26	No Health Consequences or Impact	A13	Communication or Transmission Problem		
16653	2025		47270	Cardiac resynchronization therapy implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A0711	Interrogation Problem		
									A1207	Incomplete or Inadequate Connection		
							F2301	Additional Device Required	A0711	Interrogation Problem		
									A1207	Incomplete or Inadequate Connection		
16654	2025	All other reportable incidents	47265	Dual-chamber implantable pacemaker, rate-responsive	CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A070909	Oversensing		
							F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A070909	Oversensing		
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A070909	Oversensing		
							F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A070909	Oversensing		

					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A070909	Oversensing		
							F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A070909	Oversensing		
16655	2025	All other reportable incidents	36106	Implantable pacemaker analyser, intraoperative	CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A1107	Data Problem		
							F26	No Health Consequences or Impact	A1107	Data Problem		
16656	2025	All other reportable incidents	36106	Implantable pacemaker analyser, intraoperative	CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A11	Computer Software Problem		
									A2204	Inadequate User Interface		
							F26	No Health Consequences or Impact	A11	Computer Software Problem		
									A2204	Inadequate User Interface		
16657	2025	All other reportable incidents	47265	Dual-chamber implantable pacemaker, rate-responsive	CT1685	Cardiac pacemakers	E060109	Tachycardia	A0712	Pacing Problem		
							F11	Non-Serious Injury/ Illness/ Impairment	A0712	Pacing Problem		
					CT257	Implantable cardiac pacemakers	E060109	Tachycardia	A0712	Pacing Problem		
							F11	Non-Serious Injury/ Illness/ Impairment	A0712	Pacing Problem		
					CT2866	Implantable cardiac electrical stimulators	E060109	Tachycardia	A0712	Pacing Problem		
							F11	Non-Serious Injury/ Illness/ Impairment	A0712	Pacing Problem		
16658	2025	All other reportable incidents	37265	Dual-chamber implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A070909	Oversensing		
							F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A070909	Oversensing		
16659	2025	All other reportable incidents	47270	Cardiac resynchronization therapy implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	23/12/2024	23/12/2024
									A0908	Incorrect, Inadequate or Imprecise Result or Readings	23/12/2024	23/12/2024
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	23/12/2024	23/12/2024
									A0908	Incorrect, Inadequate or Imprecise Result or Readings	23/12/2024	23/12/2024
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	23/12/2024	23/12/2024
									A0908	Incorrect, Inadequate or Imprecise Result or Readings	23/12/2024	23/12/2024
16660	2025		35852	Single-chamber implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E0602	Cardiac Arrest	A07	Electrical /Electronic Property Problem		
									A071301	Failure to Deliver Shock/Stimulation		
							F02	Death	A07	Electrical /Electronic Property Problem		
									A071301	Failure to Deliver Shock/Stimulation		
16661	2025	All other reportable incidents	47270	Cardiac resynchronization therapy implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E210401	Shock from Patient Lead(s)	A070909	Oversensing	04/11/2024	04/11/2024
									A071302	Inappropriate/Inadequate Shock/Stimulation	04/11/2024	04/11/2024
							F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A070909	Oversensing	04/11/2024	04/11/2024
									A071302	Inappropriate/Inadequate Shock/Stimulation	04/11/2024	04/11/2024
16662	2025	All other reportable incidents	17882	Physiologic-monitoring defibrillation system	CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070603	Failure to Charge		
							F27	Problem Identified During Non-Clinical Procedure	A070603	Failure to Charge		
16663	2025	All other reportable incidents	60789	Intracardiac pacemaker	CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A051201	Device Dislodged or Dislocated	21/02/2025	21/02/2025
									A070102	High Capture Threshold	21/02/2025	21/02/2025
									A070910	Undersensing	21/02/2025	21/02/2025
									A090811	Unable to Obtain Readings	21/02/2025	21/02/2025
							F08	Hospitalization or Prolonged Hospitalization	A051201	Device Dislodged or Dislocated	21/02/2025	21/02/2025
									A070102	High Capture Threshold	21/02/2025	21/02/2025
									A070910	Undersensing	21/02/2025	21/02/2025
									A090811	Unable to Obtain Readings	21/02/2025	21/02/2025
							F12	Serious Injury/ Illness/ Impairment	A051201	Device Dislodged or Dislocated	21/02/2025	21/02/2025
									A070102	High Capture Threshold	21/02/2025	21/02/2025
									A070910	Undersensing	21/02/2025	21/02/2025
									A090811	Unable to Obtain Readings	21/02/2025	21/02/2025
							F1905	Device Revision or Replacement	A051201	Device Dislodged or Dislocated	21/02/2025	21/02/2025
									A070102	High Capture Threshold	21/02/2025	21/02/2025
									A070910	Undersensing	21/02/2025	21/02/2025
									A090811	Unable to Obtain Readings	21/02/2025	21/02/2025
							F2307	Reprogramming of Device	A051201	Device Dislodged or Dislocated	21/02/2025	21/02/2025
									A070102	High Capture Threshold	21/02/2025	21/02/2025
									A070910	Undersensing	21/02/2025	21/02/2025
									A090811	Unable to Obtain Readings	21/02/2025	21/02/2025
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A051201	Device Dislodged or Dislocated	21/02/2025	21/02/2025
									A070102	High Capture Threshold	21/02/2025	21/02/2025
									A070910	Undersensing	21/02/2025	21/02/2025
									A090811	Unable to Obtain Readings	21/02/2025	21/02/2025
							F08	Hospitalization or Prolonged Hospitalization	A051201	Device Dislodged or Dislocated	21/02/2025	21/02/2025
									A070102	High Capture Threshold	21/02/2025	21/02/2025
									A070910	Undersensing	21/02/2025	21/02/2025
									A090811	Unable to Obtain Readings	21/02/2025	21/02/2025
							F12	Serious Injury/ Illness/ Impairment	A051201	Device Dislodged or Dislocated	21/02/2025	21/02/2025
									A070102	High Capture Threshold	21/02/2025	21/02/2025
									A070910	Undersensing	21/02/2025	21/02/2025
									A090811	Unable to Obtain Readings	21/02/2025	21/02/2025
							F1905	Device Revision or Replacement	A051201	Device Dislodged or Dislocated	21/02/2025	21/02/2025
									A070102	High Capture Threshold	21/02/2025	21/02/2025
									A070910	Undersensing	21/02/2025	21/02/2025
									A090811	Unable to Obtain Readings	21/02/2025	21/02/2025
							F2307	Reprogramming of Device	A051201	Device Dislodged or Dislocated	21/02/2025	21/02/2025
									A070102	High Capture Threshold	21/02/2025	21/02/2025
									A070910	Undersensing	21/02/2025	21/02/2025
									A090811	Unable to Obtain Readings	21/02/2025	21/02/2025
					CT2818	Catheterization/Cannulation/Fluid path devices	E2403	No Clinical Signs, Symptoms or Conditions	A051201	Device Dislodged or Dislocated	21/02/2025	21/02/2025
									A070102	High Capture Threshold	21/02/2025	21/02/2025
									A070910	Undersensing	21/02/2025	21/02/2025
									A090811	Unable to Obtain Readings	21/02/2025	21/02/2025
							F08	Hospitalization or Prolonged Hospitalization	A051201	Device Dislodged or Dislocated	21/02/2025	21/02/2025
									A070102	High Capture Threshold	21/02/2025	21/02/2025
									A070910	Undersensing	21/02/2025	21/02/2025
									A090811	Unable to Obtain Readings	21/02/2025	21/02/2025
							F12	Serious Injury/ Illness/ Impairment	A051201	Device Dislodged or Dislocated	21/02/2025	21/02/2025
									A070102	High Capture Threshold	21/02/2025	21/02/2025
									A070910	Undersensing	21/02/2025	21/02/2025
									A090811	Unable to Obtain Readings	21/02/2025	21/02/2025
							F1905	Device Revision or Replacement	A051201	Device Dislodged or Dislocated	21/02/2025	21/02/2025

								A070102	High Capture Threshold	21/02/2025	21/02/2025
								A070910	Undersensing	21/02/2025	21/02/2025
								A090811	Unable to Obtain Readings	21/02/2025	21/02/2025
						F2307	Reprogramming of Device	A051201	Device Dislodged or Dislocated	21/02/2025	21/02/2025
								A070102	High Capture Threshold	21/02/2025	21/02/2025
								A070910	Undersensing	21/02/2025	21/02/2025
								A090811	Unable to Obtain Readings	21/02/2025	21/02/2025
					CT2866	Implantable cardiac electrical stimulators	E2403	A051201	Device Dislodged or Dislocated	21/02/2025	21/02/2025
								A070102	High Capture Threshold	21/02/2025	21/02/2025
								A070910	Undersensing	21/02/2025	21/02/2025
								A090811	Unable to Obtain Readings	21/02/2025	21/02/2025
							F08	A051201	Device Dislodged or Dislocated	21/02/2025	21/02/2025
								A070102	High Capture Threshold	21/02/2025	21/02/2025
								A070910	Undersensing	21/02/2025	21/02/2025
								A090811	Unable to Obtain Readings	21/02/2025	21/02/2025
							F12	A051201	Device Dislodged or Dislocated	21/02/2025	21/02/2025
								A070102	High Capture Threshold	21/02/2025	21/02/2025
								A070910	Undersensing	21/02/2025	21/02/2025
								A090811	Unable to Obtain Readings	21/02/2025	21/02/2025
							F1905	A051201	Device Dislodged or Dislocated	21/02/2025	21/02/2025
								A070102	High Capture Threshold	21/02/2025	21/02/2025
								A070910	Undersensing	21/02/2025	21/02/2025
								A090811	Unable to Obtain Readings	21/02/2025	21/02/2025
							F2307	A051201	Device Dislodged or Dislocated	21/02/2025	21/02/2025
								A070102	High Capture Threshold	21/02/2025	21/02/2025
								A070910	Undersensing	21/02/2025	21/02/2025
								A090811	Unable to Obtain Readings	21/02/2025	21/02/2025
16664	2025	All other reportable incidents	37265	Dual-chamber implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E2403	A070909	Oversensing		
								A0712	Pacing Problem		
								A090801	Signal Artifact/Noise		
								A1301	Failure to Read Input Signal		
							F2203	A070909	Oversensing		
								A0712	Pacing Problem		
								A090801	Signal Artifact/Noise		
								A1301	Failure to Read Input Signal		
16665	2025	All other reportable incidents	47270	Cardiac resynchronization therapy implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E2403	A0709	Device Sensing Problem		
								A090801	Signal Artifact/Noise		
								A150202	Malposition of Device		
							F08	A0709	Device Sensing Problem		
								A090801	Signal Artifact/Noise		
								A150202	Malposition of Device		
							F1904	A0709	Device Sensing Problem		
								A090801	Signal Artifact/Noise		
								A150202	Malposition of Device		
							F2203	A0709	Device Sensing Problem		
								A090801	Signal Artifact/Noise		
								A150202	Malposition of Device		
16666	2025	All other reportable incidents	60789	Intracardiac pacemaker	CT1685	Cardiac pacemakers	E2403	A070101	Failure to Capture		
								A070102	High Capture Threshold		
								A070504	Premature Discharge of Battery		
							F12	A070101	Failure to Capture		
								A070102	High Capture Threshold		
								A070504	Premature Discharge of Battery		
							F2307	A070101	Failure to Capture		
								A070102	High Capture Threshold		
								A070504	Premature Discharge of Battery		
					CT257	Implantable cardiac pacemakers	E2403	A070101	Failure to Capture		
								A070102	High Capture Threshold		
								A070504	Premature Discharge of Battery		
							F12	A070101	Failure to Capture		
								A070102	High Capture Threshold		
								A070504	Premature Discharge of Battery		
							F2307	A070101	Failure to Capture		
								A070102	High Capture Threshold		
								A070504	Premature Discharge of Battery		
					CT2818	Catheterization/Cannulation/Fluid path devices	E2403	A070101	Failure to Capture		
								A070102	High Capture Threshold		
								A070504	Premature Discharge of Battery		
							F12	A070101	Failure to Capture		
								A070102	High Capture Threshold		
								A070504	Premature Discharge of Battery		
							F2307	A070101	Failure to Capture		
								A070102	High Capture Threshold		
								A070504	Premature Discharge of Battery		
					CT2866	Implantable cardiac electrical stimulators	E2403	A070101	Failure to Capture		
								A070102	High Capture Threshold		
								A070504	Premature Discharge of Battery		
							F12	A070101	Failure to Capture		
								A070102	High Capture Threshold		
								A070504	Premature Discharge of Battery		
							F2307	A070101	Failure to Capture		
								A070102	High Capture Threshold		
								A070504	Premature Discharge of Battery		
16667	2025	All other reportable incidents	35852	Single-chamber implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E2403	A070909	Oversensing		
								A072201	High Impedance		
							F10	A070909	Oversensing		
								A072201	High Impedance		
16668	2025	All other reportable incidents	47270	Cardiac resynchronization therapy implantable	CT2866	Implantable cardiac electrical stimulators	E2403	A070504	Premature Discharge of Battery	20/12/2024	20/12/2024

				defibrillator			F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	20/12/2024	20/12/2024
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	20/12/2024	20/12/2024
16669	2025	All other reportable incidents	17882	Physiologic-monitoring defibrillation system	CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A0719	Unexpected Shutdown		
							F27	Problem Identified During Non-Clinical Procedure	A0719	Unexpected Shutdown		
16670	2025	All other reportable incidents	47267	Single-chamber implantable pacemaker, rate-responsive	CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
16671	2025	All other reportable incidents	47265	Dual-chamber implantable pacemaker, rate-responsive	CT1685	Cardiac pacemakers	E0112	Dizziness	A070101	Failure to Capture	12/02/2025	12/02/2025
									A0908	Incorrect, Inadequate or Imprecise Result or Readings	12/02/2025	12/02/2025
							E011903	Syncope/Fainting	A070101	Failure to Capture	12/02/2025	12/02/2025
									A0908	Incorrect, Inadequate or Imprecise Result or Readings	12/02/2025	12/02/2025
							E2007	Fall	A070101	Failure to Capture	12/02/2025	12/02/2025
									A0908	Incorrect, Inadequate or Imprecise Result or Readings	12/02/2025	12/02/2025
							F08	Hospitalization or Prolonged Hospitalization	A070101	Failure to Capture	12/02/2025	12/02/2025
									A0908	Incorrect, Inadequate or Imprecise Result or Readings	12/02/2025	12/02/2025
							F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A070101	Failure to Capture	12/02/2025	12/02/2025
									A0908	Incorrect, Inadequate or Imprecise Result or Readings	12/02/2025	12/02/2025
							F1905	Device Revision or Replacement	A070101	Failure to Capture	12/02/2025	12/02/2025
									A0908	Incorrect, Inadequate or Imprecise Result or Readings	12/02/2025	12/02/2025
					CT257	Implantable cardiac pacemakers	E0112	Dizziness	A070101	Failure to Capture	12/02/2025	12/02/2025
									A0908	Incorrect, Inadequate or Imprecise Result or Readings	12/02/2025	12/02/2025
							E011903	Syncope/Fainting	A070101	Failure to Capture	12/02/2025	12/02/2025
									A0908	Incorrect, Inadequate or Imprecise Result or Readings	12/02/2025	12/02/2025
							E2007	Fall	A070101	Failure to Capture	12/02/2025	12/02/2025
									A0908	Incorrect, Inadequate or Imprecise Result or Readings	12/02/2025	12/02/2025
							F08	Hospitalization or Prolonged Hospitalization	A070101	Failure to Capture	12/02/2025	12/02/2025
									A0908	Incorrect, Inadequate or Imprecise Result or Readings	12/02/2025	12/02/2025
							F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A070101	Failure to Capture	12/02/2025	12/02/2025
									A0908	Incorrect, Inadequate or Imprecise Result or Readings	12/02/2025	12/02/2025
							F1905	Device Revision or Replacement	A070101	Failure to Capture	12/02/2025	12/02/2025
									A0908	Incorrect, Inadequate or Imprecise Result or Readings	12/02/2025	12/02/2025
					CT2866	Implantable cardiac electrical stimulators	E0112	Dizziness	A070101	Failure to Capture	12/02/2025	12/02/2025
									A0908	Incorrect, Inadequate or Imprecise Result or Readings	12/02/2025	12/02/2025
							E011903	Syncope/Fainting	A070101	Failure to Capture	12/02/2025	12/02/2025
									A0908	Incorrect, Inadequate or Imprecise Result or Readings	12/02/2025	12/02/2025
							E2007	Fall	A070101	Failure to Capture	12/02/2025	12/02/2025
									A0908	Incorrect, Inadequate or Imprecise Result or Readings	12/02/2025	12/02/2025
							F08	Hospitalization or Prolonged Hospitalization	A070101	Failure to Capture	12/02/2025	12/02/2025
									A0908	Incorrect, Inadequate or Imprecise Result or Readings	12/02/2025	12/02/2025
							F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A070101	Failure to Capture	12/02/2025	12/02/2025
									A0908	Incorrect, Inadequate or Imprecise Result or Readings	12/02/2025	12/02/2025
							F1905	Device Revision or Replacement	A070101	Failure to Capture	12/02/2025	12/02/2025
									A0908	Incorrect, Inadequate or Imprecise Result or Readings	12/02/2025	12/02/2025
16672	2025	All other reportable incidents	36106	Implantable pacemaker analyser, intraoperative	CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A090201	Device Displays Incorrect Message		
							F26	No Health Consequences or Impact	A090201	Device Displays Incorrect Message		
16673	2025	All other reportable incidents	60789	Intracardiac pacemaker	CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT2818	Catheterization/Cannulation/Fluid path devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
16674	2025	All other reportable incidents	37265	Dual-chamber implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E210401	Shock from Patient Lead(s)	A0709	Device Sensing Problem		
									A0712	Pacing Problem		
									A071302	Inappropriate/Inadequate Shock/Stimulation		
							F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A0709	Device Sensing Problem		
									A0712	Pacing Problem		
									A071302	Inappropriate/Inadequate Shock/Stimulation		
							F1203	Life Threatening Illness or Injury	A0709	Device Sensing Problem		
									A0712	Pacing Problem		
									A071302	Inappropriate/Inadequate Shock/Stimulation		
							F2307	Reprogramming of Device	A0709	Device Sensing Problem		
									A0712	Pacing Problem		
									A071302	Inappropriate/Inadequate Shock/Stimulation		
16675	2025	All other reportable incidents	47267	Single-chamber implantable pacemaker, rate-responsive	CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	17/02/2025	17/02/2025
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	17/02/2025	17/02/2025
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	17/02/2025	17/02/2025
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	17/02/2025	17/02/2025
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	17/02/2025	17/02/2025
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	17/02/2025	17/02/2025

					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	17/02/2025	17/02/2025
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	17/02/2025	17/02/2025
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	17/02/2025	17/02/2025
16676	2025	All other reportable incidents	35852	Single-chamber implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	15/03/2025	15/03/2025
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	15/03/2025	15/03/2025
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	15/03/2025	15/03/2025
16677	2025	All other reportable incidents	47265	Dual-chamber implantable pacemaker, rate-responsive	CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A071102	Failure to Interrogate		
							F26	No Health Consequences or Impact	A071102	Failure to Interrogate		
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A071102	Failure to Interrogate		
							F26	No Health Consequences or Impact	A071102	Failure to Interrogate		
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A071102	Failure to Interrogate		
							F26	No Health Consequences or Impact	A071102	Failure to Interrogate		
16678	2025	All other reportable incidents	35852	Single-chamber implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem	04/10/2024	04/10/2024
									A070504	Premature Discharge of Battery	04/10/2024	04/10/2024
							F08	Hospitalization or Prolonged Hospitalization	A0705	Battery Problem	04/10/2024	04/10/2024
									A070504	Premature Discharge of Battery	04/10/2024	04/10/2024
							F1905	Device Revision or Replacement	A0705	Battery Problem	04/10/2024	04/10/2024
									A070504	Premature Discharge of Battery	04/10/2024	04/10/2024
16679	2025	All other reportable incidents	47267	Single-chamber implantable pacemaker, rate-responsive	CT1685	Cardiac pacemakers	E1906	Unspecified Infection	A24	No Apparent Device Problem	04/02/2025	04/02/2025
							E231501	Purulent Discharge	A24	No Apparent Device Problem	04/02/2025	04/02/2025
							F08	Hospitalization or Prolonged Hospitalization	A24	No Apparent Device Problem	04/02/2025	04/02/2025
							F1903	Device Explantation	A24	No Apparent Device Problem	04/02/2025	04/02/2025
							F1907	More Complex Surgery	A24	No Apparent Device Problem	04/02/2025	04/02/2025
							F2203	Imaging Required	A24	No Apparent Device Problem	04/02/2025	04/02/2025
							F2303	Medication Required	A24	No Apparent Device Problem	04/02/2025	04/02/2025
					CT257	Implantable cardiac pacemakers	E1906	Unspecified Infection	A24	No Apparent Device Problem	04/02/2025	04/02/2025
							E231501	Purulent Discharge	A24	No Apparent Device Problem	04/02/2025	04/02/2025
							F08	Hospitalization or Prolonged Hospitalization	A24	No Apparent Device Problem	04/02/2025	04/02/2025
							F1903	Device Explantation	A24	No Apparent Device Problem	04/02/2025	04/02/2025
							F1907	More Complex Surgery	A24	No Apparent Device Problem	04/02/2025	04/02/2025
							F2203	Imaging Required	A24	No Apparent Device Problem	04/02/2025	04/02/2025
							F2303	Medication Required	A24	No Apparent Device Problem	04/02/2025	04/02/2025
					CT2866	Implantable cardiac electrical stimulators	E1906	Unspecified Infection	A24	No Apparent Device Problem	04/02/2025	04/02/2025
							E231501	Purulent Discharge	A24	No Apparent Device Problem	04/02/2025	04/02/2025
							F08	Hospitalization or Prolonged Hospitalization	A24	No Apparent Device Problem	04/02/2025	04/02/2025
							F1903	Device Explantation	A24	No Apparent Device Problem	04/02/2025	04/02/2025
							F1907	More Complex Surgery	A24	No Apparent Device Problem	04/02/2025	04/02/2025
							F2203	Imaging Required	A24	No Apparent Device Problem	04/02/2025	04/02/2025
							F2303	Medication Required	A24	No Apparent Device Problem	04/02/2025	04/02/2025
16680	2025	All other reportable incidents	47265	Dual-chamber implantable pacemaker, rate-responsive	CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A071102	Failure to Interrogate		
							F26	No Health Consequences or Impact	A071102	Failure to Interrogate		
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A071102	Failure to Interrogate		
							F26	No Health Consequences or Impact	A071102	Failure to Interrogate		
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A071102	Failure to Interrogate		
							F26	No Health Consequences or Impact	A071102	Failure to Interrogate		
16681	2025	All other reportable incidents	47270	Cardiac resynchronization therapy implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E210401	Shock from Patient Lead(s)	A040101	Fracture	14/02/2025	14/02/2025
									A070909	Oversensing	14/02/2025	14/02/2025
									A071302	Inappropriate/Inadequate Shock/Stimulation	14/02/2025	14/02/2025
									A090801	Signal Artifact/Noise	14/02/2025	14/02/2025
							F08	Hospitalization or Prolonged Hospitalization	A040101	Fracture	14/02/2025	14/02/2025
									A070909	Oversensing	14/02/2025	14/02/2025
									A071302	Inappropriate/Inadequate Shock/Stimulation	14/02/2025	14/02/2025
									A090801	Signal Artifact/Noise	14/02/2025	14/02/2025
							F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A040101	Fracture	14/02/2025	14/02/2025
									A070909	Oversensing	14/02/2025	14/02/2025
									A071302	Inappropriate/Inadequate Shock/Stimulation	14/02/2025	14/02/2025
									A090801	Signal Artifact/Noise	14/02/2025	14/02/2025
							F1905	Device Revision or Replacement	A040101	Fracture	14/02/2025	14/02/2025
									A070909	Oversensing	14/02/2025	14/02/2025
									A071302	Inappropriate/Inadequate Shock/Stimulation	14/02/2025	14/02/2025
									A090801	Signal Artifact/Noise	14/02/2025	14/02/2025
							F2203	Imaging Required	A040101	Fracture	14/02/2025	14/02/2025
									A070909	Oversensing	14/02/2025	14/02/2025
									A071302	Inappropriate/Inadequate Shock/Stimulation	14/02/2025	14/02/2025
									A090801	Signal Artifact/Noise	14/02/2025	14/02/2025
16682	2025		47265	Dual-chamber implantable pacemaker, rate-responsive	CT1685	Cardiac pacemakers	E0112	Dizziness	A0705	Battery Problem		
							E2007	Fall	A0705	Battery Problem		
							F08	Hospitalization or Prolonged Hospitalization	A0705	Battery Problem		
							F11	Non-Serious Injury/ Illness/ Impairment	A0705	Battery Problem		
					CT257	Implantable cardiac pacemakers	E0112	Dizziness	A0705	Battery Problem		
							E2007	Fall	A0705	Battery Problem		
							F08	Hospitalization or Prolonged Hospitalization	A0705	Battery Problem		
							F11	Non-Serious Injury/ Illness/ Impairment	A0705	Battery Problem		
					CT2866	Implantable cardiac electrical stimulators	E0112	Dizziness	A0705	Battery Problem		
							E2007	Fall	A0705	Battery Problem		
							F08	Hospitalization or Prolonged Hospitalization	A0705	Battery Problem		
							F11	Non-Serious Injury/ Illness/ Impairment	A0705	Battery Problem		
16683	2025	All other reportable incidents	47270	Cardiac resynchronization therapy implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A0722	Impedance Problem		
									A072201	High Impedance		
									A1206	Misconnection		
							F08	Hospitalization or Prolonged Hospitalization	A0722	Impedance Problem		
									A072201	High Impedance		
									A1206	Misconnection		
							F12	Serious Injury/ Illness/ Impairment	A0722	Impedance Problem		
									A072201	High Impedance		
									A1206	Misconnection		
							F1905	Device Revision or Replacement	A0722	Impedance Problem		
									A072201	High Impedance		

									A1206	Misconnection		
							F2307	Reprogramming of Device	A0722	Impedance Problem		
									A072201	High Impedance		
									A1206	Misconnection		
16684	2025	All other reportable incidents	17882	Physiologic-monitoring defibrillation system	CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A0719	Unexpected Shutdown		
							F26	No Health Consequences or Impact	A0719	Unexpected Shutdown		
16685	2025	All other reportable incidents			CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070102	High Capture Threshold	20/03/2024	20/03/2024
							F1908	Prolonged Surgery	A070102	High Capture Threshold	20/03/2024	20/03/2024
16686	2025	All other reportable incidents	17882	Physiologic-monitoring defibrillation system	CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A07	Electrical /Electronic Property Problem		
									A070803	Failure to Power Up		
							F27	Problem Identified During Non-Clinical Procedure	A07	Electrical /Electronic Property Problem		
									A070803	Failure to Power Up		
16687	2025		47265	Dual-chamber implantable pacemaker, rate-responsive	CT1685	Cardiac pacemakers	E060109	Tachycardia	A071102	Failure to Interrogate		
									A071206	Pacing Asynchronously		
							F08	Hospitalization or Prolonged Hospitalization	A071102	Failure to Interrogate		
									A071206	Pacing Asynchronously		
							F1905	Device Revision or Replacement	A071102	Failure to Interrogate		
									A071206	Pacing Asynchronously		
					CT257	Implantable cardiac pacemakers	E060109	Tachycardia	A071102	Failure to Interrogate		
									A071206	Pacing Asynchronously		
							F08	Hospitalization or Prolonged Hospitalization	A071102	Failure to Interrogate		
									A071206	Pacing Asynchronously		
							F1905	Device Revision or Replacement	A071102	Failure to Interrogate		
									A071206	Pacing Asynchronously		
					CT2866	Implantable cardiac electrical stimulators	E060109	Tachycardia	A071102	Failure to Interrogate		
									A071206	Pacing Asynchronously		
							F08	Hospitalization or Prolonged Hospitalization	A071102	Failure to Interrogate		
									A071206	Pacing Asynchronously		
							F1905	Device Revision or Replacement	A071102	Failure to Interrogate		
									A071206	Pacing Asynchronously		
16688	2025	All other reportable incidents	47270	Cardiac resynchronization therapy implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
									A0908	Incorrect, Inadequate or Imprecise Result or Readings		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
									A0908	Incorrect, Inadequate or Imprecise Result or Readings		
16689	2025	All other reportable incidents	47265	Dual-chamber implantable pacemaker, rate-responsive			E0623	Unspecified Heart Problem	A0908	Incorrect, Inadequate or Imprecise Result or Readings	05/03/2025	05/03/2025
							F08	Hospitalization or Prolonged Hospitalization	A0908	Incorrect, Inadequate or Imprecise Result or Readings	05/03/2025	05/03/2025
							F1905	Device Revision or Replacement	A0908	Incorrect, Inadequate or Imprecise Result or Readings	05/03/2025	05/03/2025
16690	2025	All other reportable incidents	47270	Cardiac resynchronization therapy implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A072201	High Impedance	27/02/2025	27/02/2025
							F08	Hospitalization or Prolonged Hospitalization	A072201	High Impedance	27/02/2025	27/02/2025
							F1905	Device Revision or Replacement	A072201	High Impedance	27/02/2025	27/02/2025
16691	2025	All other reportable incidents	37265	Dual-chamber implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A070909	Oversensing		
							F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A070909	Oversensing		
16692	2025	All other reportable incidents			CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070910	Undersensing		
									A0712	Pacing Problem		
							F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A070910	Undersensing		
									A0712	Pacing Problem		
							F15	Recognized Device or Procedural Complication	A070910	Undersensing		
									A0712	Pacing Problem		
16693	2025	All other reportable incidents	47265	Dual-chamber implantable pacemaker, rate-responsive	CT1685	Cardiac pacemakers	E0717	Dyspnea	A0705	Battery Problem		
									A071102	Failure to Interrogate		
									A13	Communication or Transmission Problem		
									A1605	Reset Problem		
							F11	Non-Serious Injury/ Illness/ Impairment	A0705	Battery Problem		
									A071102	Failure to Interrogate		
									A13	Communication or Transmission Problem		
									A1605	Reset Problem		
							F2307	Reprogramming of Device	A0705	Battery Problem		
									A071102	Failure to Interrogate		
									A13	Communication or Transmission Problem		
									A1605	Reset Problem		
					CT257	Implantable cardiac pacemakers	E0717	Dyspnea	A0705	Battery Problem		
									A071102	Failure to Interrogate		
									A13	Communication or Transmission Problem		
									A1605	Reset Problem		
							F11	Non-Serious Injury/ Illness/ Impairment	A0705	Battery Problem		
									A071102	Failure to Interrogate		
									A13	Communication or Transmission Problem		
									A1605	Reset Problem		
							F2307	Reprogramming of Device	A0705	Battery Problem		
									A071102	Failure to Interrogate		
									A13	Communication or Transmission Problem		
									A1605	Reset Problem		
					CT2866	Implantable cardiac electrical stimulators	E0717	Dyspnea	A0705	Battery Problem		
									A071102	Failure to Interrogate		
									A13	Communication or Transmission Problem		
									A1605	Reset Problem		
							F11	Non-Serious Injury/ Illness/ Impairment	A0705	Battery Problem		
									A071102	Failure to Interrogate		
									A13	Communication or Transmission Problem		
									A1605	Reset Problem		
							F2307	Reprogramming of Device	A0705	Battery Problem		
									A071102	Failure to Interrogate		
									A13	Communication or Transmission Problem		
									A1605	Reset Problem		

16694	2025	All other reportable incidents	47267	Single-chamber implantable pacemaker, rate-responsive	CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A070909	Oversensing		
									A0712	Pacing Problem		
									A12	Connection Problem		
							F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A070909	Oversensing		
									A0712	Pacing Problem		
									A12	Connection Problem		
							F1001	Absence of Treatment	A070909	Oversensing		
									A0712	Pacing Problem		
									A12	Connection Problem		
							F2203	Imaging Required	A070909	Oversensing		
									A0712	Pacing Problem		
									A12	Connection Problem		
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A070909	Oversensing		
									A0712	Pacing Problem		
									A12	Connection Problem		
							F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A070909	Oversensing		
									A0712	Pacing Problem		
									A12	Connection Problem		
							F1001	Absence of Treatment	A070909	Oversensing		
									A0712	Pacing Problem		
									A12	Connection Problem		
							F2203	Imaging Required	A070909	Oversensing		
									A0712	Pacing Problem		
									A12	Connection Problem		
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A070909	Oversensing		
									A0712	Pacing Problem		
									A12	Connection Problem		
							F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A070909	Oversensing		
									A0712	Pacing Problem		
									A12	Connection Problem		
							F1001	Absence of Treatment	A070909	Oversensing		
									A0712	Pacing Problem		
									A12	Connection Problem		
							F2203	Imaging Required	A070909	Oversensing		
									A0712	Pacing Problem		
									A12	Connection Problem		
16695	2025	All other reportable incidents	35852	Single-chamber implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	21/05/2025	21/05/2025
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	21/05/2025	21/05/2025
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	21/05/2025	21/05/2025
16696	2025	All other reportable incidents	17882	Physiologic-monitoring defibrillation system	CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A0719	Unexpected Shutdown		
							F27	Problem Identified During Non-Clinical Procedure	A0719	Unexpected Shutdown		
16697	2025	All other reportable incidents	37265	Dual-chamber implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E0112	Dizziness	A070101	Failure to Capture		
							F26	No Health Consequences or Impact	A070101	Failure to Capture		
16698	2025	All other reportable incidents	37265	Dual-chamber implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A070909	Oversensing		
									A072001	Electromagnetic Interference		
							F26	No Health Consequences or Impact	A070909	Oversensing		
									A072001	Electromagnetic Interference		
16699	2025	All other reportable incidents	36106	Implantable pacemaker analyser, intraoperative	CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A1104	Computer Operating System Problem		
							F26	No Health Consequences or Impact	A1104	Computer Operating System Problem		
16700	2025	All other reportable incidents	47265	Dual-chamber implantable pacemaker, rate-responsive	CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A13	Communication or Transmission Problem		
									A1605	Reset Problem		
							F26	No Health Consequences or Impact	A13	Communication or Transmission Problem		
									A1605	Reset Problem		
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A13	Communication or Transmission Problem		
									A1605	Reset Problem		
							F26	No Health Consequences or Impact	A13	Communication or Transmission Problem		
									A1605	Reset Problem		
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A13	Communication or Transmission Problem		
									A1605	Reset Problem		
							F26	No Health Consequences or Impact	A13	Communication or Transmission Problem		
									A1605	Reset Problem		
16701	2025	All other reportable incidents	47265	Dual-chamber implantable pacemaker, rate-responsive	CT1685	Cardiac pacemakers	E060101	Asystole	A0712	Pacing Problem	20/02/2025	20/02/2025
									A1107	Data Problem	20/02/2025	20/02/2025
							F08	Hospitalization or Prolonged Hospitalization	A0712	Pacing Problem	20/02/2025	20/02/2025
									A1107	Data Problem	20/02/2025	20/02/2025
							F1001	Absence of Treatment	A0712	Pacing Problem	20/02/2025	20/02/2025
									A1107	Data Problem	20/02/2025	20/02/2025
							F1905	Device Revision or Replacement	A0712	Pacing Problem	20/02/2025	20/02/2025
									A1107	Data Problem	20/02/2025	20/02/2025
					CT257	Implantable cardiac pacemakers	E060101	Asystole	A0712	Pacing Problem	20/02/2025	20/02/2025
									A1107	Data Problem	20/02/2025	20/02/2025
							F08	Hospitalization or Prolonged Hospitalization	A0712	Pacing Problem	20/02/2025	20/02/2025
									A1107	Data Problem	20/02/2025	20/02/2025
							F1001	Absence of Treatment	A0712	Pacing Problem	20/02/2025	20/02/2025
									A1107	Data Problem	20/02/2025	20/02/2025
							F1905	Device Revision or Replacement	A0712	Pacing Problem	20/02/2025	20/02/2025
									A1107	Data Problem	20/02/2025	20/02/2025
					CT2866	Implantable cardiac electrical stimulators	E060101	Asystole	A0712	Pacing Problem	20/02/2025	20/02/2025
									A1107	Data Problem	20/02/2025	20/02/2025
							F08	Hospitalization or Prolonged Hospitalization	A0712	Pacing Problem	20/02/2025	20/02/2025
									A1107	Data Problem	20/02/2025	20/02/2025
							F1001	Absence of Treatment	A0712	Pacing Problem	20/02/2025	20/02/2025
									A1107	Data Problem	20/02/2025	20/02/2025
							F1905	Device Revision or Replacement	A0712	Pacing Problem	20/02/2025	20/02/2025
									A1107	Data Problem	20/02/2025	20/02/2025
16702	2025	All other reportable incidents	47265	Dual-chamber implantable pacemaker, rate-responsive	CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A0908	Incorrect, Inadequate or Imprecise Result or Readings	04/03/2025	04/03/2025

							F08	Hospitalization or Prolonged Hospitalization	A0908	Incorrect, Inadequate or Imprecise Result or Readings	04/03/2025	04/03/2025
							F1905	Device Revision or Replacement	A0908	Incorrect, Inadequate or Imprecise Result or Readings	04/03/2025	04/03/2025
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A0908	Incorrect, Inadequate or Imprecise Result or Readings	04/03/2025	04/03/2025
							F08	Hospitalization or Prolonged Hospitalization	A0908	Incorrect, Inadequate or Imprecise Result or Readings	04/03/2025	04/03/2025
							F1905	Device Revision or Replacement	A0908	Incorrect, Inadequate or Imprecise Result or Readings	04/03/2025	04/03/2025
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A0908	Incorrect, Inadequate or Imprecise Result or Readings	04/03/2025	04/03/2025
							F08	Hospitalization or Prolonged Hospitalization	A0908	Incorrect, Inadequate or Imprecise Result or Readings	04/03/2025	04/03/2025
							F1905	Device Revision or Replacement	A0908	Incorrect, Inadequate or Imprecise Result or Readings	04/03/2025	04/03/2025
16703	2025	All other reportable incidents	47270	Cardiac resynchronization therapy implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A0712	Pacing Problem		
							F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A0712	Pacing Problem		
16704	2025	All other reportable incidents	47270	Cardiac resynchronization therapy implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
									A0908	Incorrect, Inadequate or Imprecise Result or Readings		
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery		
									A0908	Incorrect, Inadequate or Imprecise Result or Readings		
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery		
									A0908	Incorrect, Inadequate or Imprecise Result or Readings		
16705	2025	All other reportable incidents	47270	Cardiac resynchronization therapy implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	11/04/2025	11/04/2025
									A0908	Incorrect, Inadequate or Imprecise Result or Readings	11/04/2025	11/04/2025
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery	11/04/2025	11/04/2025
									A0908	Incorrect, Inadequate or Imprecise Result or Readings	11/04/2025	11/04/2025
16706	2025	All other reportable incidents	47267	Single-chamber implantable pacemaker, rate-responsive	CT1685	Cardiac pacemakers	E011903	Syncope/Fainting	A070504	Premature Discharge of Battery	28/02/2025	28/02/2025
									A071102	Failure to Interrogate	28/02/2025	28/02/2025
									A071204	No Pacing	28/02/2025	28/02/2025
							E0601	Arrhythmia	A070504	Premature Discharge of Battery	28/02/2025	28/02/2025
									A071102	Failure to Interrogate	28/02/2025	28/02/2025
									A071204	No Pacing	28/02/2025	28/02/2025
							F1901	Additional Surgery	A070504	Premature Discharge of Battery	28/02/2025	28/02/2025
									A071102	Failure to Interrogate	28/02/2025	28/02/2025
									A071204	No Pacing	28/02/2025	28/02/2025
					CT257	Implantable cardiac pacemakers	E011903	Syncope/Fainting	A070504	Premature Discharge of Battery	28/02/2025	28/02/2025
									A071102	Failure to Interrogate	28/02/2025	28/02/2025
									A071204	No Pacing	28/02/2025	28/02/2025
							E0601	Arrhythmia	A070504	Premature Discharge of Battery	28/02/2025	28/02/2025
									A071102	Failure to Interrogate	28/02/2025	28/02/2025
									A071204	No Pacing	28/02/2025	28/02/2025
							F1901	Additional Surgery	A070504	Premature Discharge of Battery	28/02/2025	28/02/2025
									A071102	Failure to Interrogate	28/02/2025	28/02/2025
									A071204	No Pacing	28/02/2025	28/02/2025
					CT2866	Implantable cardiac electrical stimulators	E011903	Syncope/Fainting	A070504	Premature Discharge of Battery	28/02/2025	28/02/2025
									A071102	Failure to Interrogate	28/02/2025	28/02/2025
									A071204	No Pacing	28/02/2025	28/02/2025
							E0601	Arrhythmia	A070504	Premature Discharge of Battery	28/02/2025	28/02/2025
									A071102	Failure to Interrogate	28/02/2025	28/02/2025
									A071204	No Pacing	28/02/2025	28/02/2025
							F1901	Additional Surgery	A070504	Premature Discharge of Battery	28/02/2025	28/02/2025
									A071102	Failure to Interrogate	28/02/2025	28/02/2025
									A071204	No Pacing	28/02/2025	28/02/2025
16707	2025	All other reportable incidents	47263	Cardiac resynchronization therapy implantable pacemaker	CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A160105	False Alarm		
							F26	No Health Consequences or Impact	A160105	False Alarm		
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A160105	False Alarm		
							F26	No Health Consequences or Impact	A160105	False Alarm		
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A160105	False Alarm		
							F26	No Health Consequences or Impact	A160105	False Alarm		
16708	2025	All other reportable incidents	35852	Single-chamber implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E060110	Ventricular Fibrillation	A0709	Device Sensing Problem		
									A0712	Pacing Problem		
							F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A0709	Device Sensing Problem		
									A0712	Pacing Problem		
							F1203	Life Threatening Illness or Injury	A0709	Device Sensing Problem		
									A0712	Pacing Problem		
							F2307	Reprogramming of Device	A0709	Device Sensing Problem		
									A0712	Pacing Problem		
16709	2025		17882	Physiologic-monitoring defibrillation system	CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem		
							F27	Problem Identified During Non-Clinical Procedure	A0705	Battery Problem		
16710	2025	All other reportable incidents	17882	Physiologic-monitoring defibrillation system	CT1684	Cardiac pacemakers and associated devices			A07	Electrical /Electronic Property Problem		
16711	2025	All other reportable incidents	47205	Cardiac pulse generator programmer	CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A071102	Failure to Interrogate		
									A1107	Data Problem		
									A1305	Wireless Communication Problem		
							F26	No Health Consequences or Impact	A071102	Failure to Interrogate		
									A1107	Data Problem		
									A1305	Wireless Communication Problem		
16712	2025	All other reportable incidents	47270	Cardiac resynchronization therapy implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem	13/08/2025	13/08/2025
									A070504	Premature Discharge of Battery	13/08/2025	13/08/2025
							F08	Hospitalization or Prolonged Hospitalization	A0705	Battery Problem	13/08/2025	13/08/2025
									A070504	Premature Discharge of Battery	13/08/2025	13/08/2025
							F1905	Device Revision or Replacement	A0705	Battery Problem	13/08/2025	13/08/2025
									A070504	Premature Discharge of Battery	13/08/2025	13/08/2025
16713	2025	All other reportable incidents	47270	Cardiac resynchronization therapy implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A090801	Signal Artifact/Noise		
							F26	No Health Consequences or Impact	A090801	Signal Artifact/Noise		

16714	2025	All other reportable incidents	47265	Dual-chamber implantable pacemaker, rate-responsive	CT1685	Cardiac pacemakers	E011902	Presyncope	A070504	Premature Discharge of Battery	03/03/2025	03/03/2025
									A0908	Incorrect, Inadequate or Imprecise Result or Readings	03/03/2025	03/03/2025
							E0623	Unspecified Heart Problem	A070504	Premature Discharge of Battery	03/03/2025	03/03/2025
									A0908	Incorrect, Inadequate or Imprecise Result or Readings	03/03/2025	03/03/2025
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	03/03/2025	03/03/2025
									A0908	Incorrect, Inadequate or Imprecise Result or Readings	03/03/2025	03/03/2025
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	03/03/2025	03/03/2025
									A0908	Incorrect, Inadequate or Imprecise Result or Readings	03/03/2025	03/03/2025
					CT257	Implantable cardiac pacemakers	E011902	Presyncope	A070504	Premature Discharge of Battery	03/03/2025	03/03/2025
									A0908	Incorrect, Inadequate or Imprecise Result or Readings	03/03/2025	03/03/2025
							E0623	Unspecified Heart Problem	A070504	Premature Discharge of Battery	03/03/2025	03/03/2025
									A0908	Incorrect, Inadequate or Imprecise Result or Readings	03/03/2025	03/03/2025
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	03/03/2025	03/03/2025
									A0908	Incorrect, Inadequate or Imprecise Result or Readings	03/03/2025	03/03/2025
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	03/03/2025	03/03/2025
									A0908	Incorrect, Inadequate or Imprecise Result or Readings	03/03/2025	03/03/2025
					CT2866	Implantable cardiac electrical stimulators	E011902	Presyncope	A070504	Premature Discharge of Battery	03/03/2025	03/03/2025
									A0908	Incorrect, Inadequate or Imprecise Result or Readings	03/03/2025	03/03/2025
							E0623	Unspecified Heart Problem	A070504	Premature Discharge of Battery	03/03/2025	03/03/2025
									A0908	Incorrect, Inadequate or Imprecise Result or Readings	03/03/2025	03/03/2025
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	03/03/2025	03/03/2025
									A0908	Incorrect, Inadequate or Imprecise Result or Readings	03/03/2025	03/03/2025
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	03/03/2025	03/03/2025
									A0908	Incorrect, Inadequate or Imprecise Result or Readings	03/03/2025	03/03/2025
16715	2025	All other reportable incidents	47263	Cardiac resynchronization therapy implantable pacemaker	CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A0712	Pacing Problem		
							F26	No Health Consequences or Impact	A0712	Pacing Problem		
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A0712	Pacing Problem		
							F26	No Health Consequences or Impact	A0712	Pacing Problem		
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A0712	Pacing Problem		
							F26	No Health Consequences or Impact	A0712	Pacing Problem		
16716	2025	All other reportable incidents	47265	Dual-chamber implantable pacemaker, rate-responsive	CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A071102	Failure to Interrogate		
							F26	No Health Consequences or Impact	A071102	Failure to Interrogate		
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A071102	Failure to Interrogate		
							F26	No Health Consequences or Impact	A071102	Failure to Interrogate		
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A071102	Failure to Interrogate		
							F26	No Health Consequences or Impact	A071102	Failure to Interrogate		
16717	2025		47270	Cardiac resynchronization therapy implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A070909	Oversensing		
									A071204	No Pacing		
							F23	Unexpected Medical Intervention	A070909	Oversensing		
									A071204	No Pacing		
16718	2025	All other reportable incidents	35852	Single-chamber implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E060110	Ventricular Fibrillation	A0709	Device Sensing Problem		
									A0713	Defibrillation/Stimulation Problem		
							F05	Delay to Treatment/ Therapy	A0709	Device Sensing Problem		
									A0713	Defibrillation/Stimulation Problem		
							F1203	Life Threatening Illness or Injury	A0709	Device Sensing Problem		
									A0713	Defibrillation/Stimulation Problem		
							F2307	Reprogramming of Device	A0709	Device Sensing Problem		
									A0713	Defibrillation/Stimulation Problem		
16719	2025	All other reportable incidents	47265	Dual-chamber implantable pacemaker, rate-responsive	CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	20/02/2025	20/02/2025
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	20/02/2025	20/02/2025
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	20/02/2025	20/02/2025
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	20/02/2025	20/02/2025
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	20/02/2025	20/02/2025
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	20/02/2025	20/02/2025
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	20/02/2025	20/02/2025
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	20/02/2025	20/02/2025
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	20/02/2025	20/02/2025
16720	2025	All other reportable incidents	47270	Cardiac resynchronization therapy implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A0711	Interrogation Problem	21/05/2024	21/05/2024
							F26	No Health Consequences or Impact	A0711	Interrogation Problem	21/05/2024	21/05/2024
16721	2025	All other reportable incidents	47263	Cardiac resynchronization therapy implantable pacemaker	CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A072201	High Impedance		
									A090801	Signal Artifact/Noise		
							F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A072201	High Impedance		
									A090801	Signal Artifact/Noise		
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A072201	High Impedance		
									A090801	Signal Artifact/Noise		
							F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A072201	High Impedance		
									A090801	Signal Artifact/Noise		
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A072201	High Impedance		
									A090801	Signal Artifact/Noise		
							F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A072201	High Impedance		
									A090801	Signal Artifact/Noise		
16722	2025	All other reportable incidents	47270	Cardiac resynchronization therapy implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	31/03/2025	31/03/2025
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	31/03/2025	31/03/2025
							F12	Serious Injury/ Illness/ Impairment	A070504	Premature Discharge of Battery	31/03/2025	31/03/2025
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	31/03/2025	31/03/2025
16723	2025	All other reportable incidents	17882	Physiologic-monitoring defibrillation system	CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A07	Electrical /Electronic Property Problem		
									A070802	Intermittent Loss of Power		
							F26	No Health Consequences or Impact	A07	Electrical /Electronic Property Problem		
									A070802	Intermittent Loss of Power		

16724	2025	All other reportable incidents	47270	Cardiac resynchronization therapy implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	23/04/2025	23/04/2025
									A0908	Incorrect, Inadequate or Imprecise Result or Readings	23/04/2025	23/04/2025
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	23/04/2025	23/04/2025
									A0908	Incorrect, Inadequate or Imprecise Result or Readings	23/04/2025	23/04/2025
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	23/04/2025	23/04/2025
									A0908	Incorrect, Inadequate or Imprecise Result or Readings	23/04/2025	23/04/2025
16725	2025	All other reportable incidents	35852	Single-chamber implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem	17/03/2025	17/03/2025
									A070504	Premature Discharge of Battery	17/03/2025	17/03/2025
							F08	Hospitalization or Prolonged Hospitalization	A0705	Battery Problem	17/03/2025	17/03/2025
									A070504	Premature Discharge of Battery	17/03/2025	17/03/2025
							F12	Serious Injury/ Illness/ Impairment	A0705	Battery Problem	17/03/2025	17/03/2025
									A070504	Premature Discharge of Battery	17/03/2025	17/03/2025
							F1905	Device Revision or Replacement	A0705	Battery Problem	17/03/2025	17/03/2025
									A070504	Premature Discharge of Battery	17/03/2025	17/03/2025
16726	2025	All other reportable incidents	47270	Cardiac resynchronization therapy implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
									A090801	Signal Artifact/Noise		
									A1301	Failure to Read Input Signal		
							F2203	Imaging Required	A070504	Premature Discharge of Battery		
									A090801	Signal Artifact/Noise		
									A1301	Failure to Read Input Signal		
16727	2025		47265	Dual-chamber implantable pacemaker, rate-responsive	CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A1305	Wireless Communication Problem		
							F26	No Health Consequences or Impact	A1305	Wireless Communication Problem		
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A1305	Wireless Communication Problem		
							F26	No Health Consequences or Impact	A1305	Wireless Communication Problem		
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A1305	Wireless Communication Problem		
							F26	No Health Consequences or Impact	A1305	Wireless Communication Problem		
16728	2025	All other reportable incidents	35852	Single-chamber implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	06/03/2025	06/03/2025
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	06/03/2025	06/03/2025
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	06/03/2025	06/03/2025
16729	2025	All other reportable incidents	47265	Dual-chamber implantable pacemaker, rate-responsive	CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
16730	2025	All other reportable incidents	35852	Single-chamber implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	24/04/2025	24/04/2025
									A0908	Incorrect, Inadequate or Imprecise Result or Readings	24/04/2025	24/04/2025
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	24/04/2025	24/04/2025
									A0908	Incorrect, Inadequate or Imprecise Result or Readings	24/04/2025	24/04/2025
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	24/04/2025	24/04/2025
									A0908	Incorrect, Inadequate or Imprecise Result or Readings	24/04/2025	24/04/2025
16731	2025	All other reportable incidents	47263	Cardiac resynchronization therapy implantable pacemaker	CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A0908	Incorrect, Inadequate or Imprecise Result or Readings		
							F26	No Health Consequences or Impact	A0908	Incorrect, Inadequate or Imprecise Result or Readings		
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A0908	Incorrect, Inadequate or Imprecise Result or Readings		
							F26	No Health Consequences or Impact	A0908	Incorrect, Inadequate or Imprecise Result or Readings		
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A0908	Incorrect, Inadequate or Imprecise Result or Readings		
							F26	No Health Consequences or Impact	A0908	Incorrect, Inadequate or Imprecise Result or Readings		
16732	2025	All other reportable incidents	60789	Intracardiac pacemaker	CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A0701	Capturing Problem	06/03/2025	06/03/2025
							F26	No Health Consequences or Impact	A0701	Capturing Problem	06/03/2025	06/03/2025
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A0701	Capturing Problem	06/03/2025	06/03/2025
							F26	No Health Consequences or Impact	A0701	Capturing Problem	06/03/2025	06/03/2025
					CT2818	Catheterization/Cannulation/Fluid path devices	E2403	No Clinical Signs, Symptoms or Conditions	A0701	Capturing Problem	06/03/2025	06/03/2025
							F26	No Health Consequences or Impact	A0701	Capturing Problem	06/03/2025	06/03/2025
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A0701	Capturing Problem	06/03/2025	06/03/2025
							F26	No Health Consequences or Impact	A0701	Capturing Problem	06/03/2025	06/03/2025
16733	2025	All other reportable incidents	36106	Implantable pacemaker analyser, intraoperative	CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A1104	Computer Operating System Problem		
							F26	No Health Consequences or Impact	A1104	Computer Operating System Problem		
16734	2025	All other reportable incidents	47270	Cardiac resynchronization therapy implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A070909	Oversensing		
									A090801	Signal Artifact/Noise		
									A150202	Malposition of Device		
							F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A070909	Oversensing		
									A090801	Signal Artifact/Noise		
									A150202	Malposition of Device		
16735	2025		57968	Cardiovascular device introducer, steerable	CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A25	No Apparent Adverse Event		
							F26	No Health Consequences or Impact	A25	No Apparent Adverse Event		
16736	2025	All other reportable incidents	60789	Intracardiac pacemaker	CT1685	Cardiac pacemakers	E0604	Cardiac Perforation	A24	No Apparent Device Problem		
							E0619	Pericardial Effusion	A24	No Apparent Device Problem		
							F12	Serious Injury/ Illness/ Impairment	A24	No Apparent Device Problem		
					CT257	Implantable cardiac pacemakers	E0604	Cardiac Perforation	A24	No Apparent Device Problem		
							E0619	Pericardial Effusion	A24	No Apparent Device Problem		
							F12	Serious Injury/ Illness/ Impairment	A24	No Apparent Device Problem		
					CT2818	Catheterization/Cannulation/Fluid path devices	E0604	Cardiac Perforation	A24	No Apparent Device Problem		
							E0619	Pericardial Effusion	A24	No Apparent Device Problem		

					CT2866	Implantable cardiac electrical stimulators	F12	Serious Injury/ Illness/ Impairment	A24	No Apparent Device Problem		
							E0604	Cardiac Perforation	A24	No Apparent Device Problem		
							E0619	Pericardial Effusion	A24	No Apparent Device Problem		
							F12	Serious Injury/ Illness/ Impairment	A24	No Apparent Device Problem		
16737	2025	All other reportable incidents	57968	Cardiovascular device introducer, steerable	CT1684	Cardiac pacemakers and associated devices	E0604	Cardiac Perforation	A24	No Apparent Device Problem		
							E0619	Pericardial Effusion	A24	No Apparent Device Problem		
							F12	Serious Injury/ Illness/ Impairment	A24	No Apparent Device Problem		
16738	2025		60789	Intracardiac pacemaker	CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A25	No Apparent Adverse Event		
							F26	No Health Consequences or Impact	A25	No Apparent Adverse Event		
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A25	No Apparent Adverse Event		
							F26	No Health Consequences or Impact	A25	No Apparent Adverse Event		
					CT2818	Catheterization/Cannulation/Fluid path devices	E2403	No Clinical Signs, Symptoms or Conditions	A25	No Apparent Adverse Event		
							F26	No Health Consequences or Impact	A25	No Apparent Adverse Event		
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A25	No Apparent Adverse Event		
							F26	No Health Consequences or Impact	A25	No Apparent Adverse Event		
16739	2025	All other reportable incidents	47265	Dual-chamber implantable pacemaker, rate-responsive	CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
16740	2025	Death	17882	Physiologic-monitoring defibrillation system	CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A071301	Failure to Deliver Shock/Stimulation		
							F02	Death	A071301	Failure to Deliver Shock/Stimulation		
16741	2025	All other reportable incidents	47270	Cardiac resynchronization therapy implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	12/05/2025	12/05/2025
									A0908	Incorrect, Inadequate or Imprecise Result or Readings	12/05/2025	12/05/2025
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	12/05/2025	12/05/2025
									A0908	Incorrect, Inadequate or Imprecise Result or Readings	12/05/2025	12/05/2025
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	12/05/2025	12/05/2025
									A0908	Incorrect, Inadequate or Imprecise Result or Readings	12/05/2025	12/05/2025
16742	2025	All other reportable incidents	47270	Cardiac resynchronization therapy implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A0908	Incorrect, Inadequate or Imprecise Result or Readings		
							F26	No Health Consequences or Impact	A0908	Incorrect, Inadequate or Imprecise Result or Readings		
16743	2025	All other reportable incidents	47270	Cardiac resynchronization therapy implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A070909	Oversensing		
									A0712	Pacing Problem		
							F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A070909	Oversensing		
									A0712	Pacing Problem		
16744	2025		47265	Dual-chamber implantable pacemaker, rate-responsive	CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	06/03/2025	06/03/2025
							F1901	Additional Surgery	A070504	Premature Discharge of Battery	06/03/2025	06/03/2025
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	06/03/2025	06/03/2025
							F1901	Additional Surgery	A070504	Premature Discharge of Battery	06/03/2025	06/03/2025
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	06/03/2025	06/03/2025
							F1901	Additional Surgery	A070504	Premature Discharge of Battery	06/03/2025	06/03/2025
16745	2025	All other reportable incidents	37265	Dual-chamber implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E210401	Shock from Patient Lead(s)	A0712	Pacing Problem		
									A071302	Inappropriate/Inadequate Shock/Stimulation		
							F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A0712	Pacing Problem		
									A071302	Inappropriate/Inadequate Shock/Stimulation		
16746	2025	All other reportable incidents	47270	Cardiac resynchronization therapy implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A072202	Low Impedance		
							F26	No Health Consequences or Impact	A072202	Low Impedance		
16747	2025	All other reportable incidents	47265	Dual-chamber implantable pacemaker, rate-responsive	CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem	18/01/2024	18/01/2024
									A070504	Premature Discharge of Battery	18/01/2024	18/01/2024
							F08	Hospitalization or Prolonged Hospitalization	A0705	Battery Problem	18/01/2024	18/01/2024
									A070504	Premature Discharge of Battery	18/01/2024	18/01/2024
							F1905	Device Revision or Replacement	A0705	Battery Problem	18/01/2024	18/01/2024
									A070504	Premature Discharge of Battery	18/01/2024	18/01/2024
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem	18/01/2024	18/01/2024
									A070504	Premature Discharge of Battery	18/01/2024	18/01/2024
							F08	Hospitalization or Prolonged Hospitalization	A0705	Battery Problem	18/01/2024	18/01/2024
									A070504	Premature Discharge of Battery	18/01/2024	18/01/2024
							F1905	Device Revision or Replacement	A0705	Battery Problem	18/01/2024	18/01/2024
									A070504	Premature Discharge of Battery	18/01/2024	18/01/2024
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem	18/01/2024	18/01/2024
									A070504	Premature Discharge of Battery	18/01/2024	18/01/2024
							F08	Hospitalization or Prolonged Hospitalization	A0705	Battery Problem	18/01/2024	18/01/2024
									A070504	Premature Discharge of Battery	18/01/2024	18/01/2024
							F1905	Device Revision or Replacement	A0705	Battery Problem	18/01/2024	18/01/2024
									A070504	Premature Discharge of Battery	18/01/2024	18/01/2024
16748	2025	All other reportable incidents	47265	Dual-chamber implantable pacemaker, rate-responsive	CT1685	Cardiac pacemakers	E0112	Dizziness	A070909	Oversensing	21/02/2025	21/02/2025
									A0712	Pacing Problem	21/02/2025	21/02/2025
									A0908	Incorrect, Inadequate or Imprecise Result or Readings	21/02/2025	21/02/2025
							E011903	Syncope/Fainting	A070909	Oversensing	21/02/2025	21/02/2025
									A0712	Pacing Problem	21/02/2025	21/02/2025
									A0908	Incorrect, Inadequate or Imprecise Result or Readings	21/02/2025	21/02/2025
							E1610	Head Injury	A070909	Oversensing	21/02/2025	21/02/2025
									A0712	Pacing Problem	21/02/2025	21/02/2025
									A0908	Incorrect, Inadequate or Imprecise Result or Readings	21/02/2025	21/02/2025
							E2007	Fall	A070909	Oversensing	21/02/2025	21/02/2025
									A0712	Pacing Problem	21/02/2025	21/02/2025
									A0908	Incorrect, Inadequate or Imprecise Result or Readings	21/02/2025	21/02/2025
							F08	Hospitalization or Prolonged Hospitalization	A070909	Oversensing	21/02/2025	21/02/2025
									A0712	Pacing Problem	21/02/2025	21/02/2025
									A0908	Incorrect, Inadequate or Imprecise Result or Readings	21/02/2025	21/02/2025

							F1001	Absence of Treatment	A070909	Oversensing	21/02/2025	21/02/2025
									A0712	Pacing Problem	21/02/2025	21/02/2025
									A0908	Incorrect, Inadequate or Imprecise Result or Readings	21/02/2025	21/02/2025
							F1905	Device Revision or Replacement	A070909	Oversensing	21/02/2025	21/02/2025
									A0712	Pacing Problem	21/02/2025	21/02/2025
									A0908	Incorrect, Inadequate or Imprecise Result or Readings	21/02/2025	21/02/2025
					CT257	Implantable cardiac pacemakers	E0112	Dizziness	A070909	Oversensing	21/02/2025	21/02/2025
									A0712	Pacing Problem	21/02/2025	21/02/2025
									A0908	Incorrect, Inadequate or Imprecise Result or Readings	21/02/2025	21/02/2025
							E011903	Syncope/Fainting	A070909	Oversensing	21/02/2025	21/02/2025
									A0712	Pacing Problem	21/02/2025	21/02/2025
									A0908	Incorrect, Inadequate or Imprecise Result or Readings	21/02/2025	21/02/2025
							E1610	Head Injury	A070909	Oversensing	21/02/2025	21/02/2025
									A0712	Pacing Problem	21/02/2025	21/02/2025
									A0908	Incorrect, Inadequate or Imprecise Result or Readings	21/02/2025	21/02/2025
							E2007	Fall	A070909	Oversensing	21/02/2025	21/02/2025
									A0712	Pacing Problem	21/02/2025	21/02/2025
									A0908	Incorrect, Inadequate or Imprecise Result or Readings	21/02/2025	21/02/2025
							F08	Hospitalization or Prolonged Hospitalization	A070909	Oversensing	21/02/2025	21/02/2025
									A0712	Pacing Problem	21/02/2025	21/02/2025
									A0908	Incorrect, Inadequate or Imprecise Result or Readings	21/02/2025	21/02/2025
							F1001	Absence of Treatment	A070909	Oversensing	21/02/2025	21/02/2025
									A0712	Pacing Problem	21/02/2025	21/02/2025
									A0908	Incorrect, Inadequate or Imprecise Result or Readings	21/02/2025	21/02/2025
							F1905	Device Revision or Replacement	A070909	Oversensing	21/02/2025	21/02/2025
									A0712	Pacing Problem	21/02/2025	21/02/2025
									A0908	Incorrect, Inadequate or Imprecise Result or Readings	21/02/2025	21/02/2025
					CT2866	Implantable cardiac electrical stimulators	E0112	Dizziness	A070909	Oversensing	21/02/2025	21/02/2025
									A0712	Pacing Problem	21/02/2025	21/02/2025
									A0908	Incorrect, Inadequate or Imprecise Result or Readings	21/02/2025	21/02/2025
							E011903	Syncope/Fainting	A070909	Oversensing	21/02/2025	21/02/2025
									A0712	Pacing Problem	21/02/2025	21/02/2025
									A0908	Incorrect, Inadequate or Imprecise Result or Readings	21/02/2025	21/02/2025
							E1610	Head Injury	A070909	Oversensing	21/02/2025	21/02/2025
									A0712	Pacing Problem	21/02/2025	21/02/2025
									A0908	Incorrect, Inadequate or Imprecise Result or Readings	21/02/2025	21/02/2025
							E2007	Fall	A070909	Oversensing	21/02/2025	21/02/2025
									A0712	Pacing Problem	21/02/2025	21/02/2025
									A0908	Incorrect, Inadequate or Imprecise Result or Readings	21/02/2025	21/02/2025
							F08	Hospitalization or Prolonged Hospitalization	A070909	Oversensing	21/02/2025	21/02/2025
									A0712	Pacing Problem	21/02/2025	21/02/2025
									A0908	Incorrect, Inadequate or Imprecise Result or Readings	21/02/2025	21/02/2025
							F1001	Absence of Treatment	A070909	Oversensing	21/02/2025	21/02/2025
									A0712	Pacing Problem	21/02/2025	21/02/2025
									A0908	Incorrect, Inadequate or Imprecise Result or Readings	21/02/2025	21/02/2025
							F1905	Device Revision or Replacement	A070909	Oversensing	21/02/2025	21/02/2025
									A0712	Pacing Problem	21/02/2025	21/02/2025
									A0908	Incorrect, Inadequate or Imprecise Result or Readings	21/02/2025	21/02/2025
16749	2025	All other reportable incidents	47270	Cardiac resynchronization therapy implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A0203	Defective Device	12/12/2024	12/12/2024
									A070504	Premature Discharge of Battery	12/12/2024	12/12/2024
							F08	Hospitalization or Prolonged Hospitalization	A0203	Defective Device	12/12/2024	12/12/2024
									A070504	Premature Discharge of Battery	12/12/2024	12/12/2024
							F1905	Device Revision or Replacement	A0203	Defective Device	12/12/2024	12/12/2024
									A070504	Premature Discharge of Battery	12/12/2024	12/12/2024
16750	2025	All other reportable incidents	17882	Physiologic-monitoring defibrillation system	CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A0719	Unexpected Shutdown		
							F27	Problem Identified During Non-Clinical Procedure	A0719	Unexpected Shutdown		
16751	2025	All other reportable incidents	47267	Single-chamber implantable pacemaker, rate-responsive	CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
16752	2025	All other reportable incidents	47270	Cardiac resynchronization therapy implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
									A0908	Incorrect, Inadequate or Imprecise Result or Readings		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
									A0908	Incorrect, Inadequate or Imprecise Result or Readings		
16753	2025	All other reportable incidents			CT1684	Cardiac pacemakers and associated devices	E060101	Asystole	A070909	Oversensing		
									A0712	Pacing Problem		
							F1001	Absence of Treatment	A070909	Oversensing		
									A0712	Pacing Problem		
							F1006	Inadequate Treatment/Disease Progression	A070909	Oversensing		
									A0712	Pacing Problem		
16754	2025	All other reportable incidents	47265	Dual-chamber implantable pacemaker, rate-responsive	CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A0908	Incorrect, Inadequate or Imprecise Result or Readings	12/03/2025	12/03/2025

							F08	Hospitalization or Prolonged Hospitalization	A0908	Incorrect, Inadequate or Imprecise Result or Readings	12/03/2025	12/03/2025
							F1905	Device Revision or Replacement	A0908	Incorrect, Inadequate or Imprecise Result or Readings	12/03/2025	12/03/2025
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A0908	Incorrect, Inadequate or Imprecise Result or Readings	12/03/2025	12/03/2025
							F08	Hospitalization or Prolonged Hospitalization	A0908	Incorrect, Inadequate or Imprecise Result or Readings	12/03/2025	12/03/2025
							F1905	Device Revision or Replacement	A0908	Incorrect, Inadequate or Imprecise Result or Readings	12/03/2025	12/03/2025
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A0908	Incorrect, Inadequate or Imprecise Result or Readings	12/03/2025	12/03/2025
							F08	Hospitalization or Prolonged Hospitalization	A0908	Incorrect, Inadequate or Imprecise Result or Readings	12/03/2025	12/03/2025
							F1905	Device Revision or Replacement	A0908	Incorrect, Inadequate or Imprecise Result or Readings	12/03/2025	12/03/2025
16755	2025		17882	Physiologic-monitoring defibrillation system	CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A0713	Defibrillation/Stimulation Problem		
									A0902	Display or Visual Feedback Problem		
							F2301	Additional Device Required	A0713	Defibrillation/Stimulation Problem		
									A0902	Display or Visual Feedback Problem		
16756	2025	All other reportable incidents	47270	Cardiac resynchronization therapy implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E210401	Shock from Patient Lead(s)	A070909	Oversensing		
									A071302	Inappropriate/Inadequate Shock/Stimulation		
									A090801	Signal Artifact/Noise		
									A1301	Failure to Read Input Signal		
							F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A070909	Oversensing		
									A071302	Inappropriate/Inadequate Shock/Stimulation		
									A090801	Signal Artifact/Noise		
									A1301	Failure to Read Input Signal		
							F2203	Imaging Required	A070909	Oversensing		
									A071302	Inappropriate/Inadequate Shock/Stimulation		
									A090801	Signal Artifact/Noise		
									A1301	Failure to Read Input Signal		
16757	2025	All other reportable incidents	47270	Cardiac resynchronization therapy implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
									A0908	Incorrect, Inadequate or Imprecise Result or Readings		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
									A0908	Incorrect, Inadequate or Imprecise Result or Readings		
16758	2025	All other reportable incidents	47265	Dual-chamber implantable pacemaker, rate-responsive	CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A0908	Incorrect, Inadequate or Imprecise Result or Readings	04/04/2025	04/04/2025
							F26	No Health Consequences or Impact	A0908	Incorrect, Inadequate or Imprecise Result or Readings	04/04/2025	04/04/2025
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A0908	Incorrect, Inadequate or Imprecise Result or Readings	04/04/2025	04/04/2025
							F26	No Health Consequences or Impact	A0908	Incorrect, Inadequate or Imprecise Result or Readings	04/04/2025	04/04/2025
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A0908	Incorrect, Inadequate or Imprecise Result or Readings	04/04/2025	04/04/2025
							F26	No Health Consequences or Impact	A0908	Incorrect, Inadequate or Imprecise Result or Readings	04/04/2025	04/04/2025
16759	2025	All other reportable incidents	47263	Cardiac resynchronization therapy implantable pacemaker	CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A0104	Migration or Expulsion of Device		
							F08	Hospitalization or Prolonged Hospitalization	A0104	Migration or Expulsion of Device		
							F12	Serious Injury/ Illness/ Impairment	A0104	Migration or Expulsion of Device		
							F1904	Device Repositioning	A0104	Migration or Expulsion of Device		
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A0104	Migration or Expulsion of Device		
							F08	Hospitalization or Prolonged Hospitalization	A0104	Migration or Expulsion of Device		
							F12	Serious Injury/ Illness/ Impairment	A0104	Migration or Expulsion of Device		
							F1904	Device Repositioning	A0104	Migration or Expulsion of Device		
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A0104	Migration or Expulsion of Device		
							F08	Hospitalization or Prolonged Hospitalization	A0104	Migration or Expulsion of Device		
							F12	Serious Injury/ Illness/ Impairment	A0104	Migration or Expulsion of Device		
							F1904	Device Repositioning	A0104	Migration or Expulsion of Device		
16760	2025	All other reportable incidents	37265	Dual-chamber implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E060109	Tachycardia	A0709	Device Sensing Problem		
									A0712	Pacing Problem		
							F11	Non-Serious Injury/ Illness/ Impairment	A0709	Device Sensing Problem		
									A0712	Pacing Problem		
16761	2025	All other reportable incidents	63188	Left-ventricular wireless pacing system sensor/transmitter	CT145	Cardiovascular devices			A0104	Migration or Expulsion of Device		
					CT1577	Wireless monitoring system transmitters			A0104	Migration or Expulsion of Device		
					CT1684	Cardiac pacemakers and associated devices			A0104	Migration or Expulsion of Device		
					CT2727	Active implant devices/ancillaries			A0104	Migration or Expulsion of Device		
					CT2750	Implantable cardiac electrical stimulators and associated devices			A0104	Migration or Expulsion of Device		
					CT2885	Physiological monitoring/measuring system transmitters/receivers/transceivers			A0104	Migration or Expulsion of Device		
16762	2025	All other reportable incidents	47267	Single-chamber implantable pacemaker, rate-responsive	CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	26/02/2025	26/02/2025
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	26/02/2025	26/02/2025
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	26/02/2025	26/02/2025
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	26/02/2025	26/02/2025
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	26/02/2025	26/02/2025
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	26/02/2025	26/02/2025
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	26/02/2025	26/02/2025
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	26/02/2025	26/02/2025
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	26/02/2025	26/02/2025
16763	2025		17882	Physiologic-monitoring defibrillation system	CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A0203	Defective Device		
									A072103	Intermittent Continuity		
							F26	No Health Consequences or Impact	A0203	Defective Device		
									A072103	Intermittent Continuity		
16764	2025	All other reportable incidents	35852	Single-chamber implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A160601	Premature Elective Replacement Indicator		
							F26	No Health Consequences or Impact	A160601	Premature Elective Replacement Indicator		
16765	2025	All other reportable incidents	37265	Dual-chamber implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	27/03/2025	27/03/2025
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	27/03/2025	27/03/2025
							F12	Serious Injury/ Illness/ Impairment	A070504	Premature Discharge of Battery	27/03/2025	27/03/2025

							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	27/03/2025	27/03/2025
16766	2025		47265	Dual-chamber implantable pacemaker, rate-responsive	CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	05/03/2025	05/03/2025
							F1901	Additional Surgery	A24	No Apparent Device Problem	05/03/2025	05/03/2025
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	05/03/2025	05/03/2025
							F1901	Additional Surgery	A24	No Apparent Device Problem	05/03/2025	05/03/2025
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	05/03/2025	05/03/2025
							F1901	Additional Surgery	A24	No Apparent Device Problem	05/03/2025	05/03/2025
16767	2025	All other reportable incidents	35852	Single-chamber implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	11/07/2025	11/07/2025
									A0722	Impedance Problem	11/07/2025	11/07/2025
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	11/07/2025	11/07/2025
									A0722	Impedance Problem	11/07/2025	11/07/2025
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	11/07/2025	11/07/2025
									A0722	Impedance Problem	11/07/2025	11/07/2025
16768	2025	All other reportable incidents	17882	Physiologic-monitoring defibrillation system	CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A07	Electrical /Electronic Property Problem		
							F27	Problem Identified During Non-Clinical Procedure	A07	Electrical /Electronic Property Problem		
16769	2025		17882	Physiologic-monitoring defibrillation system	CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A01	Patient Device Interaction Problem		
									A110201	Application Program Freezes or Fails to Launch		
							F26	No Health Consequences or Impact	A01	Patient Device Interaction Problem		
									A110201	Application Program Freezes or Fails to Launch		
16770	2025	All other reportable incidents			CT1684	Cardiac pacemakers and associated devices	E1906	Unspecified Infection	A0203	Defective Device		
							E2311	Discomfort	A0203	Defective Device		
							E2330	Pain	A0203	Defective Device		
							F08	Hospitalization or Prolonged Hospitalization	A0203	Defective Device		
							F1901	Additional Surgery	A0203	Defective Device		
16771	2025	All other reportable incidents	47267	Single-chamber implantable pacemaker, rate-responsive	CT1685	Cardiac pacemakers	E0601	Arrhythmia	A071102	Failure to Interrogate	25/03/2025	25/03/2025
									A071204	No Pacing	25/03/2025	25/03/2025
									A0723	Unexpected Mode Switch	25/03/2025	25/03/2025
							F1901	Additional Surgery	A071102	Failure to Interrogate	25/03/2025	25/03/2025
									A071204	No Pacing	25/03/2025	25/03/2025
									A0723	Unexpected Mode Switch	25/03/2025	25/03/2025
					CT257	Implantable cardiac pacemakers	E0601	Arrhythmia	A071102	Failure to Interrogate	25/03/2025	25/03/2025
									A071204	No Pacing	25/03/2025	25/03/2025
									A0723	Unexpected Mode Switch	25/03/2025	25/03/2025
							F1901	Additional Surgery	A071102	Failure to Interrogate	25/03/2025	25/03/2025
									A071204	No Pacing	25/03/2025	25/03/2025
									A0723	Unexpected Mode Switch	25/03/2025	25/03/2025
					CT2866	Implantable cardiac electrical stimulators	E0601	Arrhythmia	A071102	Failure to Interrogate	25/03/2025	25/03/2025
									A071204	No Pacing	25/03/2025	25/03/2025
									A0723	Unexpected Mode Switch	25/03/2025	25/03/2025
							F1901	Additional Surgery	A071102	Failure to Interrogate	25/03/2025	25/03/2025
									A071204	No Pacing	25/03/2025	25/03/2025
									A0723	Unexpected Mode Switch	25/03/2025	25/03/2025
16772	2025	All other reportable incidents	47267	Single-chamber implantable pacemaker, rate-responsive	CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	17/03/2025	17/03/2025
							F1901	Additional Surgery	A24	No Apparent Device Problem	17/03/2025	17/03/2025
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	17/03/2025	17/03/2025
							F1901	Additional Surgery	A24	No Apparent Device Problem	17/03/2025	17/03/2025
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	17/03/2025	17/03/2025
							F1901	Additional Surgery	A24	No Apparent Device Problem	17/03/2025	17/03/2025
16773	2025	All other reportable incidents	47270	Cardiac resynchronization therapy implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A070909	Oversensing		
									A0712	Pacing Problem		
									A090801	Signal Artifact/Noise		
									A1301	Failure to Read Input Signal		
							F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A070909	Oversensing		
									A0712	Pacing Problem		
									A090801	Signal Artifact/Noise		
									A1301	Failure to Read Input Signal		
16774	2025	All other reportable incidents	17882	Physiologic-monitoring defibrillation system	CT1684	Cardiac pacemakers and associated devices	E0623	Unspecified Heart Problem	A071301	Failure to Deliver Shock/Stimulation		
							F05	Delay to Treatment/ Therapy	A071301	Failure to Deliver Shock/Stimulation		
16775	2025	All other reportable incidents	47265	Dual-chamber implantable pacemaker, rate-responsive	CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A0701	Capturing Problem		
									A090811	Unable to Obtain Readings		
							F12	Serious Injury/ Illness/ Impairment	A0701	Capturing Problem		
									A090811	Unable to Obtain Readings		
							F2307	Reprogramming of Device	A0701	Capturing Problem		
									A090811	Unable to Obtain Readings		
							F26	No Health Consequences or Impact	A0701	Capturing Problem		
									A090811	Unable to Obtain Readings		
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A0701	Capturing Problem		
									A090811	Unable to Obtain Readings		
							F12	Serious Injury/ Illness/ Impairment	A0701	Capturing Problem		
									A090811	Unable to Obtain Readings		
							F2307	Reprogramming of Device	A0701	Capturing Problem		
									A090811	Unable to Obtain Readings		
							F26	No Health Consequences or Impact	A0701	Capturing Problem		
									A090811	Unable to Obtain Readings		
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A0701	Capturing Problem		
									A090811	Unable to Obtain Readings		
							F12	Serious Injury/ Illness/ Impairment	A0701	Capturing Problem		
									A090811	Unable to Obtain Readings		
							F2307	Reprogramming of Device	A0701	Capturing Problem		
									A090811	Unable to Obtain Readings		
							F26	No Health Consequences or Impact	A0701	Capturing Problem		
									A090811	Unable to Obtain Readings		
16776	2025	All other reportable incidents	47265	Dual-chamber implantable pacemaker, rate-responsive	CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem		
							F26	No Health Consequences or Impact	A0705	Battery Problem		
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem		
							F26	No Health Consequences or Impact	A0705	Battery Problem		
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem		
							F26	No Health Consequences or Impact	A0705	Battery Problem		

16777	2025	All other reportable incidents	47270	Cardiac resynchronization therapy implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E060110	Ventricular Fibrillation	A070910	Undersensing		
									A0909	Unexpected Therapeutic Results		
									A1301	Failure to Read Input Signal		
							F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A070910	Undersensing		
									A0909	Unexpected Therapeutic Results		
									A1301	Failure to Read Input Signal		
							F1203	Life Threatening Illness or Injury	A070910	Undersensing		
									A0909	Unexpected Therapeutic Results		
									A1301	Failure to Read Input Signal		
16778	2025	All other reportable incidents	47270	Cardiac resynchronization therapy implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E060110	Ventricular Fibrillation	A070910	Undersensing		
									A0909	Unexpected Therapeutic Results		
									A1301	Failure to Read Input Signal		
							F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A070910	Undersensing		
									A0909	Unexpected Therapeutic Results		
									A1301	Failure to Read Input Signal		
							F1203	Life Threatening Illness or Injury	A070910	Undersensing		
									A0909	Unexpected Therapeutic Results		
									A1301	Failure to Read Input Signal		
16779	2025	All other reportable incidents	47270	Cardiac resynchronization therapy implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A072201	High Impedance		
							F26	No Health Consequences or Impact	A072201	High Impedance		
16780	2025	All other reportable incidents	47263	Cardiac resynchronization therapy implantable pacemaker	CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	01/03/2025	01/03/2025
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	01/03/2025	01/03/2025
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	01/03/2025	01/03/2025
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	01/03/2025	01/03/2025
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	01/03/2025	01/03/2025
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	01/03/2025	01/03/2025
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	01/03/2025	01/03/2025
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	01/03/2025	01/03/2025
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	01/03/2025	01/03/2025
16781	2025	All other reportable incidents	47265	Dual-chamber implantable pacemaker, rate-responsive			E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	17/03/2025	17/03/2025
									A071102	Failure to Interrogate	17/03/2025	17/03/2025
									A071204	No Pacing	17/03/2025	17/03/2025
							F1901	Additional Surgery	A070504	Premature Discharge of Battery	17/03/2025	17/03/2025
									A071102	Failure to Interrogate	17/03/2025	17/03/2025
									A071204	No Pacing	17/03/2025	17/03/2025
16782	2025	All other reportable incidents	47267	Single-chamber implantable pacemaker, rate-responsive			E2403	No Clinical Signs, Symptoms or Conditions	A0908	Incorrect, Inadequate or Imprecise Result or Readings	24/03/2025	24/03/2025
							F08	Hospitalization or Prolonged Hospitalization	A0908	Incorrect, Inadequate or Imprecise Result or Readings	24/03/2025	24/03/2025
							F1905	Device Revision or Replacement	A0908	Incorrect, Inadequate or Imprecise Result or Readings	24/03/2025	24/03/2025
16783	2025	All other reportable incidents	37265	Dual-chamber implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A0712	Pacing Problem		
							F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A0712	Pacing Problem		
16784	2025	All other reportable incidents	47265	Dual-chamber implantable pacemaker, rate-responsive	CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A2101	Device Markings/Labeling Problem		
							F26	No Health Consequences or Impact	A2101	Device Markings/Labeling Problem		
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A2101	Device Markings/Labeling Problem		
							F26	No Health Consequences or Impact	A2101	Device Markings/Labeling Problem		
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A2101	Device Markings/Labeling Problem		
							F26	No Health Consequences or Impact	A2101	Device Markings/Labeling Problem		
16785	2025	All other reportable incidents	47265	Dual-chamber implantable pacemaker, rate-responsive	CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A070909	Oversensing		
							F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A070909	Oversensing		
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A070909	Oversensing		
							F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A070909	Oversensing		
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A070909	Oversensing		
							F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A070909	Oversensing		
16786	2025	All other reportable incidents	47267	Single-chamber implantable pacemaker, rate-responsive	CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	13/11/2024	13/11/2024
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	13/11/2024	13/11/2024
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	13/11/2024	13/11/2024
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	13/11/2024	13/11/2024
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	13/11/2024	13/11/2024
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	13/11/2024	13/11/2024
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	13/11/2024	13/11/2024
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	13/11/2024	13/11/2024
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	13/11/2024	13/11/2024
16787	2025	All other reportable incidents	35852	Single-chamber implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E210401	Shock from Patient Lead(s)	A070909	Oversensing		
									A071302	Inappropriate/Inadequate Shock/Stimulation		
									A072201	High Impedance		
									A1301	Failure to Read Input Signal		
							F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A070909	Oversensing		
									A071302	Inappropriate/Inadequate Shock/Stimulation		
									A072201	High Impedance		
									A1301	Failure to Read Input Signal		
16788	2025	All other reportable incidents	47265	Dual-chamber implantable pacemaker, rate-responsive	CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A070910	Undersensing		
									A072202	Low Impedance		
									A1301	Failure to Read Input Signal		
							F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A070910	Undersensing		
									A072202	Low Impedance		
									A1301	Failure to Read Input Signal		
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A070910	Undersensing		
									A072202	Low Impedance		
									A1301	Failure to Read Input Signal		
							F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A070910	Undersensing		
									A072202	Low Impedance		
									A1301	Failure to Read Input Signal		
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A070910	Undersensing		
									A072202	Low Impedance		
									A1301	Failure to Read Input Signal		
							F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A070910	Undersensing		

									A072202	Low Impedance		
									A1301	Failure to Read Input Signal		
16789	2025	All other reportable incidents	47265	Dual-chamber implantable pacemaker, rate-responsive	CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A0908	Incorrect, Inadequate or Imprecise Result or Readings	01/03/2025	01/03/2025
							F08	Hospitalization or Prolonged Hospitalization	A0908	Incorrect, Inadequate or Imprecise Result or Readings	01/03/2025	01/03/2025
							F1905	Device Revision or Replacement	A0908	Incorrect, Inadequate or Imprecise Result or Readings	01/03/2025	01/03/2025
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A0908	Incorrect, Inadequate or Imprecise Result or Readings	01/03/2025	01/03/2025
							F08	Hospitalization or Prolonged Hospitalization	A0908	Incorrect, Inadequate or Imprecise Result or Readings	01/03/2025	01/03/2025
							F1905	Device Revision or Replacement	A0908	Incorrect, Inadequate or Imprecise Result or Readings	01/03/2025	01/03/2025
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A0908	Incorrect, Inadequate or Imprecise Result or Readings	01/03/2025	01/03/2025
							F08	Hospitalization or Prolonged Hospitalization	A0908	Incorrect, Inadequate or Imprecise Result or Readings	01/03/2025	01/03/2025
							F1905	Device Revision or Replacement	A0908	Incorrect, Inadequate or Imprecise Result or Readings	01/03/2025	01/03/2025
16790	2025	All other reportable incidents	63188	Left-ventricular wireless pacing system sensor/transmitter			E0623	Unspecified Heart Problem	A0708	Power Problem		
							F24	Insufficient Health Impact Information	A0708	Power Problem		
16791	2025	All other reportable incidents	17882	Physiologic-monitoring defibrillation system	CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A0719	Unexpected Shutdown		
							F26	No Health Consequences or Impact	A0719	Unexpected Shutdown		
16792	2025	All other reportable incidents	47205	Cardiac pulse generator programmer	CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A0902	Display or Visual Feedback Problem		
							F27	Problem Identified During Non-Clinical Procedure	A0902	Display or Visual Feedback Problem		
16793	2025	All other reportable incidents	35852	Single-chamber implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	20/06/2025	20/06/2025
									A0908	Incorrect, Inadequate or Imprecise Result or Readings	20/06/2025	20/06/2025
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	20/06/2025	20/06/2025
									A0908	Incorrect, Inadequate or Imprecise Result or Readings	20/06/2025	20/06/2025
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	20/06/2025	20/06/2025
									A0908	Incorrect, Inadequate or Imprecise Result or Readings	20/06/2025	20/06/2025
16794	2025	All other reportable incidents	17882	Physiologic-monitoring defibrillation system	CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A07	Electrical /Electronic Property Problem		
							F27	Problem Identified During Non-Clinical Procedure	A07	Electrical /Electronic Property Problem		
16795	2025	All other reportable incidents	17882	Physiologic-monitoring defibrillation system	CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A0908	Incorrect, Inadequate or Imprecise Result or Readings		
							F27	Problem Identified During Non-Clinical Procedure	A0908	Incorrect, Inadequate or Imprecise Result or Readings		
16796	2025	All other reportable incidents	35852	Single-chamber implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A072201	High Impedance		
							F26	No Health Consequences or Impact	A072201	High Impedance		
16797	2025	All other reportable incidents	17882	Physiologic-monitoring defibrillation system	CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A090201	Device Displays Incorrect Message		
							F27	Problem Identified During Non-Clinical Procedure	A090201	Device Displays Incorrect Message		
16798	2025	All other reportable incidents	36106	Implantable pacemaker analyser, intraoperative	CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A11	Computer Software Problem		
							F26	No Health Consequences or Impact	A11	Computer Software Problem		
16799	2025	All other reportable incidents	36106	Implantable pacemaker analyser, intraoperative	CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A11	Computer Software Problem		
									A1102	Application Program Problem		
									A110207	Problem with Software Installation		
							F26	No Health Consequences or Impact	A11	Computer Software Problem		
									A1102	Application Program Problem		
									A110207	Problem with Software Installation		
16800	2025	All other reportable incidents	47265	Dual-chamber implantable pacemaker, rate-responsive	CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	02/04/2025	02/04/2025
									A0908	Incorrect, Inadequate or Imprecise Result or Readings	02/04/2025	02/04/2025
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	02/04/2025	02/04/2025
									A0908	Incorrect, Inadequate or Imprecise Result or Readings	02/04/2025	02/04/2025
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	02/04/2025	02/04/2025
									A0908	Incorrect, Inadequate or Imprecise Result or Readings	02/04/2025	02/04/2025
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	02/04/2025	02/04/2025
									A0908	Incorrect, Inadequate or Imprecise Result or Readings	02/04/2025	02/04/2025
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	02/04/2025	02/04/2025
									A0908	Incorrect, Inadequate or Imprecise Result or Readings	02/04/2025	02/04/2025
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	02/04/2025	02/04/2025
									A0908	Incorrect, Inadequate or Imprecise Result or Readings	02/04/2025	02/04/2025
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	02/04/2025	02/04/2025
									A0908	Incorrect, Inadequate or Imprecise Result or Readings	02/04/2025	02/04/2025
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	02/04/2025	02/04/2025
									A0908	Incorrect, Inadequate or Imprecise Result or Readings	02/04/2025	02/04/2025
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	02/04/2025	02/04/2025
									A0908	Incorrect, Inadequate or Imprecise Result or Readings	02/04/2025	02/04/2025
16801	2025	All other reportable incidents	47267	Single-chamber implantable pacemaker, rate-responsive	CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A070101	Failure to Capture		
									A070909	Oversensing		
									A072202	Low Impedance		
									A090801	Signal Artifact/Noise		
							F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A070101	Failure to Capture		
									A070909	Oversensing		
									A072202	Low Impedance		
									A090801	Signal Artifact/Noise		
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A070101	Failure to Capture		
									A070909	Oversensing		
									A072202	Low Impedance		
									A090801	Signal Artifact/Noise		
							F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A070101	Failure to Capture		

									A070909	Oversensing		
									A072202	Low Impedance		
									A090801	Signal Artifact/Noise		
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A070101	Failure to Capture		
									A070909	Oversensing		
									A072202	Low Impedance		
									A090801	Signal Artifact/Noise		
							F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A070101	Failure to Capture		
									A070909	Oversensing		
									A072202	Low Impedance		
									A090801	Signal Artifact/Noise		
16802	2025	All other reportable incidents	17882	Physiologic-monitoring defibrillation system	CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A090202	Display Difficult to Read		
							F26	No Health Consequences or Impact	A090202	Display Difficult to Read		
16803	2025	All other reportable incidents	35224	External pacemaker, epicardial pacing	CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A05	Mechanical Problem		
							F2601	Problem Identified Before Clinical Use/Exposure	A05	Mechanical Problem		
16804	2025	All other reportable incidents	47270	Cardiac resynchronization therapy implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A05	Mechanical Problem	21/03/2025	21/03/2025
									A1206	Misconnection	21/03/2025	21/03/2025
							F26	No Health Consequences or Impact	A05	Mechanical Problem	21/03/2025	21/03/2025
									A1206	Misconnection	21/03/2025	21/03/2025
16805	2025	All other reportable incidents	47270	Cardiac resynchronization therapy implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A070909	Oversensing		
									A071204	No Pacing		
							F23	Unexpected Medical Intervention	A070909	Oversensing		
									A071204	No Pacing		
16806	2025	All other reportable incidents	47263	Cardiac resynchronization therapy implantable pacemaker	CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A0908	Incorrect, Inadequate or Imprecise Result or Readings	21/03/2025	21/03/2025
							F08	Hospitalization or Prolonged Hospitalization	A0908	Incorrect, Inadequate or Imprecise Result or Readings	21/03/2025	21/03/2025
							F1905	Device Revision or Replacement	A0908	Incorrect, Inadequate or Imprecise Result or Readings	21/03/2025	21/03/2025
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A0908	Incorrect, Inadequate or Imprecise Result or Readings	21/03/2025	21/03/2025
							F08	Hospitalization or Prolonged Hospitalization	A0908	Incorrect, Inadequate or Imprecise Result or Readings	21/03/2025	21/03/2025
							F1905	Device Revision or Replacement	A0908	Incorrect, Inadequate or Imprecise Result or Readings	21/03/2025	21/03/2025
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A0908	Incorrect, Inadequate or Imprecise Result or Readings	21/03/2025	21/03/2025
							F08	Hospitalization or Prolonged Hospitalization	A0908	Incorrect, Inadequate or Imprecise Result or Readings	21/03/2025	21/03/2025
							F1905	Device Revision or Replacement	A0908	Incorrect, Inadequate or Imprecise Result or Readings	21/03/2025	21/03/2025
16807	2025	All other reportable incidents	17882	Physiologic-monitoring defibrillation system	CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A0719	Unexpected Shutdown		
							F27	Problem Identified During Non-Clinical Procedure	A0719	Unexpected Shutdown		
16808	2025	All other reportable incidents	17882	Physiologic-monitoring defibrillation system	CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A07	Electrical /Electronic Property Problem		
							F27	Problem Identified During Non-Clinical Procedure	A07	Electrical /Electronic Property Problem		
16809	2025		37265	Dual-chamber implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A0709	Device Sensing Problem		
									A110301	Delayed Program or Algorithm Execution		
							F05	Delay to Treatment/ Therapy	A0709	Device Sensing Problem		
									A110301	Delayed Program or Algorithm Execution		
16810	2025	All other reportable incidents	47265	Dual-chamber implantable pacemaker, rate-responsive	CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A05	Mechanical Problem	19/03/2025	19/03/2025
							F26	No Health Consequences or Impact	A05	Mechanical Problem	19/03/2025	19/03/2025
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A05	Mechanical Problem	19/03/2025	19/03/2025
							F26	No Health Consequences or Impact	A05	Mechanical Problem	19/03/2025	19/03/2025
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A05	Mechanical Problem	19/03/2025	19/03/2025
							F26	No Health Consequences or Impact	A05	Mechanical Problem	19/03/2025	19/03/2025
16811	2025	All other reportable incidents	47265	Dual-chamber implantable pacemaker, rate-responsive	CT1685	Cardiac pacemakers	E0601	Arrhythmia	A071102	Failure to Interrogate	02/04/2025	02/04/2025
									A071204	No Pacing	02/04/2025	02/04/2025
							F1901	Additional Surgery	A071102	Failure to Interrogate	02/04/2025	02/04/2025
									A071204	No Pacing	02/04/2025	02/04/2025
					CT257	Implantable cardiac pacemakers	E0601	Arrhythmia	A071102	Failure to Interrogate	02/04/2025	02/04/2025
									A071204	No Pacing	02/04/2025	02/04/2025
							F1901	Additional Surgery	A071102	Failure to Interrogate	02/04/2025	02/04/2025
									A071204	No Pacing	02/04/2025	02/04/2025
					CT2866	Implantable cardiac electrical stimulators	E0601	Arrhythmia	A071102	Failure to Interrogate	02/04/2025	02/04/2025
									A071204	No Pacing	02/04/2025	02/04/2025
							F1901	Additional Surgery	A071102	Failure to Interrogate	02/04/2025	02/04/2025
									A071204	No Pacing	02/04/2025	02/04/2025
16812	2025	All other reportable incidents	47265	Dual-chamber implantable pacemaker, rate-responsive	CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A0709	Device Sensing Problem		
									A0712	Pacing Problem		
							F2307	Reprogramming of Device	A0709	Device Sensing Problem		
									A0712	Pacing Problem		
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A0709	Device Sensing Problem		
									A0712	Pacing Problem		
							F2307	Reprogramming of Device	A0709	Device Sensing Problem		
									A0712	Pacing Problem		
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A0709	Device Sensing Problem		
									A0712	Pacing Problem		
							F2307	Reprogramming of Device	A0709	Device Sensing Problem		
									A0712	Pacing Problem		
16813	2025	All other reportable incidents	37265	Dual-chamber implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E210401	Shock from Patient Lead(s)	A0709	Device Sensing Problem		
									A071302	Inappropriate/Inadequate Shock/Stimulation		
							F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A0709	Device Sensing Problem		
									A071302	Inappropriate/Inadequate Shock/Stimulation		
							F1203	Life Threatening Illness or Injury	A0709	Device Sensing Problem		
									A071302	Inappropriate/Inadequate Shock/Stimulation		
							F2307	Reprogramming of Device	A0709	Device Sensing Problem		
									A071302	Inappropriate/Inadequate Shock/Stimulation		
16814	2025	All other reportable incidents	47265	Dual-chamber implantable pacemaker, rate-responsive			E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem	06/06/2025	06/06/2025
									A070504	Premature Discharge of Battery	06/06/2025	06/06/2025
							F08	Hospitalization or Prolonged Hospitalization	A0705	Battery Problem	06/06/2025	06/06/2025

									A070504	Premature Discharge of Battery	06/06/2025	06/06/2025
							F1905	Device Revision or Replacement	A0705	Battery Problem	06/06/2025	06/06/2025
									A070504	Premature Discharge of Battery	06/06/2025	06/06/2025
16815	2025	All other reportable incidents	47265	Dual-chamber implantable pacemaker, rate-responsive	CT1685	Cardiac pacemakers	E2311	Discomfort	A070909	Oversensing		
									A070910	Undersensing		
									A0712	Pacing Problem		
									A072201	High Impedance		
									A090801	Signal Artifact/Noise		
									A1301	Failure to Read Input Signal		
							F1001	Absence of Treatment	A070909	Oversensing		
									A070910	Undersensing		
									A0712	Pacing Problem		
									A072201	High Impedance		
									A090801	Signal Artifact/Noise		
									A1301	Failure to Read Input Signal		
					CT257	Implantable cardiac pacemakers	E2311	Discomfort	A070909	Oversensing		
									A070910	Undersensing		
									A0712	Pacing Problem		
									A072201	High Impedance		
									A090801	Signal Artifact/Noise		
									A1301	Failure to Read Input Signal		
							F1001	Absence of Treatment	A070909	Oversensing		
									A070910	Undersensing		
									A0712	Pacing Problem		
									A072201	High Impedance		
									A090801	Signal Artifact/Noise		
									A1301	Failure to Read Input Signal		
					CT2866	Implantable cardiac electrical stimulators	E2311	Discomfort	A070909	Oversensing		
									A070910	Undersensing		
									A0712	Pacing Problem		
									A072201	High Impedance		
									A090801	Signal Artifact/Noise		
									A1301	Failure to Read Input Signal		
							F1001	Absence of Treatment	A070909	Oversensing		
									A070910	Undersensing		
									A0712	Pacing Problem		
									A072201	High Impedance		
									A090801	Signal Artifact/Noise		
									A1301	Failure to Read Input Signal		
16816	2025	All other reportable incidents	47270	Cardiac resynchronization therapy implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E060101	Asystole	A070101	Failure to Capture		
									A070910	Undersensing		
									A071201	Failure to Convert Rhythm		
									A0909	Unexpected Therapeutic Results		
							F1203	Life Threatening Illness or Injury	A070101	Failure to Capture		
									A070910	Undersensing		
									A071201	Failure to Convert Rhythm		
									A0909	Unexpected Therapeutic Results		
							F1908	Prolonged Surgery	A070101	Failure to Capture		
									A070910	Undersensing		
									A071201	Failure to Convert Rhythm		
									A0909	Unexpected Therapeutic Results		
							F2306	Resuscitation	A070101	Failure to Capture		
									A070910	Undersensing		
									A071201	Failure to Convert Rhythm		
									A0909	Unexpected Therapeutic Results		
16817	2025	All other reportable incidents	37265	Dual-chamber implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A070909	Oversensing		
									A072001	Electromagnetic Interference		
									A090801	Signal Artifact/Noise		
							F26	No Health Consequences or Impact	A070909	Oversensing		
									A072001	Electromagnetic Interference		
									A090801	Signal Artifact/Noise		
16818	2025		47265	Dual-chamber implantable pacemaker, rate-responsive	CT1685	Cardiac pacemakers	E0601	Arrhythmia	A071204	No Pacing		
							F23	Unexpected Medical Intervention	A071204	No Pacing		
					CT257	Implantable cardiac pacemakers	E0601	Arrhythmia	A071204	No Pacing		
							F23	Unexpected Medical Intervention	A071204	No Pacing		
					CT2866	Implantable cardiac electrical stimulators	E0601	Arrhythmia	A071204	No Pacing		
							F23	Unexpected Medical Intervention	A071204	No Pacing		
16819	2025	All other reportable incidents	17882	Physiologic-monitoring defibrillation system	CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A150105	Key or Button Unresponsive/not Working		
							F27	Problem Identified During Non-Clinical Procedure	A150105	Key or Button Unresponsive/not Working		
16820	2025	All other reportable incidents	47265	Dual-chamber implantable pacemaker, rate-responsive	CT1685	Cardiac pacemakers	E0717	Dyspnea	A071102	Failure to Interrogate	03/04/2025	03/04/2025
									A071204	No Pacing	03/04/2025	03/04/2025
									A0723	Unexpected Mode Switch	03/04/2025	03/04/2025
							F1901	Additional Surgery	A071102	Failure to Interrogate	03/04/2025	03/04/2025
									A071204	No Pacing	03/04/2025	03/04/2025
									A0723	Unexpected Mode Switch	03/04/2025	03/04/2025
					CT257	Implantable cardiac pacemakers	E0717	Dyspnea	A071102	Failure to Interrogate	03/04/2025	03/04/2025
									A071204	No Pacing	03/04/2025	03/04/2025
									A0723	Unexpected Mode Switch	03/04/2025	03/04/2025
							F1901	Additional Surgery	A071102	Failure to Interrogate	03/04/2025	03/04/2025
									A071204	No Pacing	03/04/2025	03/04/2025
									A0723	Unexpected Mode Switch	03/04/2025	03/04/2025
					CT2866	Implantable cardiac electrical stimulators	E0717	Dyspnea	A071102	Failure to Interrogate	03/04/2025	03/04/2025
									A071204	No Pacing	03/04/2025	03/04/2025
									A0723	Unexpected Mode Switch	03/04/2025	03/04/2025
							F1901	Additional Surgery	A071102	Failure to Interrogate	03/04/2025	03/04/2025
									A071204	No Pacing	03/04/2025	03/04/2025
									A0723	Unexpected Mode Switch	03/04/2025	03/04/2025
16821	2025	All other reportable incidents	47270	Cardiac resynchronization therapy implantable	CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A070909	Oversensing		

				defibrillator					A072201	High Impedance		
							F1001	Absence of Treatment	A070909	Oversensing		
									A072201	High Impedance		
16822	2025	All other reportable incidents	47270	Cardiac resynchronization therapy implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A072201	High Impedance		
									A1301	Failure to Read Input Signal		
							F26	No Health Consequences or Impact	A072201	High Impedance		
									A1301	Failure to Read Input Signal		
16823	2025	All other reportable incidents	37265	Dual-chamber implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A090811	Unable to Obtain Readings		
							F26	No Health Consequences or Impact	A090811	Unable to Obtain Readings		
16824	2025	All other reportable incidents	47265	Dual-chamber implantable pacemaker, rate-responsive	CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A072202	Low Impedance		
							F26	No Health Consequences or Impact	A072202	Low Impedance		
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A072202	Low Impedance		
							F26	No Health Consequences or Impact	A072202	Low Impedance		
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A072202	Low Impedance		
							F26	No Health Consequences or Impact	A072202	Low Impedance		
16825	2025	All other reportable incidents	47270	Cardiac resynchronization therapy implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	01/04/2025	01/04/2025
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	01/04/2025	01/04/2025
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	01/04/2025	01/04/2025
16826	2025	All other reportable incidents	47263	Cardiac resynchronization therapy implantable pacemaker	CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A070101	Failure to Capture		
									A0709	Device Sensing Problem		
							F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A070101	Failure to Capture		
									A0709	Device Sensing Problem		
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A070101	Failure to Capture		
									A0709	Device Sensing Problem		
							F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A070101	Failure to Capture		
									A0709	Device Sensing Problem		
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A070101	Failure to Capture		
									A0709	Device Sensing Problem		
							F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A070101	Failure to Capture		
									A0709	Device Sensing Problem		
16827	2025	All other reportable incidents	47265	Dual-chamber implantable pacemaker, rate-responsive	CT1685	Cardiac pacemakers	E0717	Dyspnea	A070909	Oversensing		
									A0712	Pacing Problem		
							F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A070909	Oversensing		
									A0712	Pacing Problem		
							F15	Recognized Device or Procedural Complication	A070909	Oversensing		
									A0712	Pacing Problem		
					CT257	Implantable cardiac pacemakers	E0717	Dyspnea	A070909	Oversensing		
									A0712	Pacing Problem		
							F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A070909	Oversensing		
									A0712	Pacing Problem		
							F15	Recognized Device or Procedural Complication	A070909	Oversensing		
									A0712	Pacing Problem		
					CT2866	Implantable cardiac electrical stimulators	E0717	Dyspnea	A070909	Oversensing		
									A0712	Pacing Problem		
							F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A070909	Oversensing		
									A0712	Pacing Problem		
							F15	Recognized Device or Procedural Complication	A070909	Oversensing		
									A0712	Pacing Problem		
16828	2025	All other reportable incidents	37265	Dual-chamber implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E2104	Electric Shock	A070909	Oversensing		
									A071302	Inappropriate/Inadequate Shock/Stimulation		
									A072201	High Impedance		
									A090801	Signal Artifact/Noise		
							E210401	Shock from Patient Lead(s)	A070909	Oversensing		
									A071302	Inappropriate/Inadequate Shock/Stimulation		
									A072201	High Impedance		
									A090801	Signal Artifact/Noise		
							F08	Hospitalization or Prolonged Hospitalization	A070909	Oversensing		
									A071302	Inappropriate/Inadequate Shock/Stimulation		
									A072201	High Impedance		
									A090801	Signal Artifact/Noise		
							F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A070909	Oversensing		
									A071302	Inappropriate/Inadequate Shock/Stimulation		
									A072201	High Impedance		
									A090801	Signal Artifact/Noise		
16829	2025	All other reportable incidents	35852	Single-chamber implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E210401	Shock from Patient Lead(s)	A0709	Device Sensing Problem		
									A071302	Inappropriate/Inadequate Shock/Stimulation		
							F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A0709	Device Sensing Problem		
									A071302	Inappropriate/Inadequate Shock/Stimulation		
							F1203	Life Threatening Illness or Injury	A0709	Device Sensing Problem		
									A071302	Inappropriate/Inadequate Shock/Stimulation		
							F2307	Reprogramming of Device	A0709	Device Sensing Problem		
									A071302	Inappropriate/Inadequate Shock/Stimulation		
16830	2025	All other reportable incidents	47265	Dual-chamber implantable pacemaker, rate-responsive	CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A070909	Oversensing		
									A072201	High Impedance		
									A090801	Signal Artifact/Noise		
							F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A070909	Oversensing		
									A072201	High Impedance		
									A090801	Signal Artifact/Noise		
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A070909	Oversensing		
									A072201	High Impedance		
									A090801	Signal Artifact/Noise		
							F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A070909	Oversensing		
									A072201	High Impedance		
									A090801	Signal Artifact/Noise		
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A070909	Oversensing		
									A072201	High Impedance		
									A090801	Signal Artifact/Noise		
							F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A070909	Oversensing		

									A072201	High Impedance		
									A090801	Signal Artifact/Noise		
16831	2025	All other reportable incidents	17882	Physiologic-monitoring defibrillation system	CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A150105	Key or Button Unresponsive/not Working		
							F27	Problem Identified During Non-Clinical Procedure	A150105	Key or Button Unresponsive/not Working		
16832	2025	All other reportable incidents	17882	Physiologic-monitoring defibrillation system	CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A150105	Key or Button Unresponsive/not Working		
							F27	Problem Identified During Non-Clinical Procedure	A150105	Key or Button Unresponsive/not Working		
16833	2025	All other reportable incidents	17882	Physiologic-monitoring defibrillation system	CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A150105	Key or Button Unresponsive/not Working		
							F27	Problem Identified During Non-Clinical Procedure	A150105	Key or Button Unresponsive/not Working		
16834	2025	All other reportable incidents	47265	Dual-chamber implantable pacemaker, rate-responsive	CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	03/04/2025	03/04/2025
									A0908	Incorrect, Inadequate or Imprecise Result or Readings	03/04/2025	03/04/2025
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	03/04/2025	03/04/2025
									A0908	Incorrect, Inadequate or Imprecise Result or Readings	03/04/2025	03/04/2025
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	03/04/2025	03/04/2025
									A0908	Incorrect, Inadequate or Imprecise Result or Readings	03/04/2025	03/04/2025
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	03/04/2025	03/04/2025
									A0908	Incorrect, Inadequate or Imprecise Result or Readings	03/04/2025	03/04/2025
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	03/04/2025	03/04/2025
									A0908	Incorrect, Inadequate or Imprecise Result or Readings	03/04/2025	03/04/2025
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	03/04/2025	03/04/2025
									A0908	Incorrect, Inadequate or Imprecise Result or Readings	03/04/2025	03/04/2025
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	03/04/2025	03/04/2025
									A0908	Incorrect, Inadequate or Imprecise Result or Readings	03/04/2025	03/04/2025
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	03/04/2025	03/04/2025
									A0908	Incorrect, Inadequate or Imprecise Result or Readings	03/04/2025	03/04/2025
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	03/04/2025	03/04/2025
									A0908	Incorrect, Inadequate or Imprecise Result or Readings	03/04/2025	03/04/2025
16835	2025	All other reportable incidents	17882	Physiologic-monitoring defibrillation system	CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A150105	Key or Button Unresponsive/not Working		
							F27	Problem Identified During Non-Clinical Procedure	A150105	Key or Button Unresponsive/not Working		
16836	2025	All other reportable incidents	17882	Physiologic-monitoring defibrillation system	CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A150105	Key or Button Unresponsive/not Working		
							F27	Problem Identified During Non-Clinical Procedure	A150105	Key or Button Unresponsive/not Working		
16837	2025	All other reportable incidents	47265	Dual-chamber implantable pacemaker, rate-responsive	CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A1605	Reset Problem		
									A160503	Inappropriate or Unexpected Reset		
							F08	Hospitalization or Prolonged Hospitalization	A1605	Reset Problem		
									A160503	Inappropriate or Unexpected Reset		
							F2307	Reprogramming of Device	A1605	Reset Problem		
									A160503	Inappropriate or Unexpected Reset		
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A1605	Reset Problem		
									A160503	Inappropriate or Unexpected Reset		
							F08	Hospitalization or Prolonged Hospitalization	A1605	Reset Problem		
									A160503	Inappropriate or Unexpected Reset		
							F2307	Reprogramming of Device	A1605	Reset Problem		
									A160503	Inappropriate or Unexpected Reset		
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A1605	Reset Problem		
									A160503	Inappropriate or Unexpected Reset		
							F08	Hospitalization or Prolonged Hospitalization	A1605	Reset Problem		
									A160503	Inappropriate or Unexpected Reset		
							F2307	Reprogramming of Device	A1605	Reset Problem		
									A160503	Inappropriate or Unexpected Reset		
16838	2025	All other reportable incidents	35852	Single-chamber implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E060110	Ventricular Fibrillation	A070910	Undersensing		
									A0713	Defibrillation/Stimulation Problem		
							F05	Delay to Treatment/ Therapy	A070910	Undersensing		
									A0713	Defibrillation/Stimulation Problem		
							F1203	Life Threatening Illness or Injury	A070910	Undersensing		
									A0713	Defibrillation/Stimulation Problem		
							F2307	Reprogramming of Device	A070910	Undersensing		
									A0713	Defibrillation/Stimulation Problem		
16839	2025	All other reportable incidents	47265	Dual-chamber implantable pacemaker, rate-responsive	CT1685	Cardiac pacemakers	E0601	Arrhythmia	A070101	Failure to Capture		
									A070909	Oversensing		
									A0712	Pacing Problem		
							F23	Unexpected Medical Intervention	A070101	Failure to Capture		
									A070909	Oversensing		
									A0712	Pacing Problem		
					CT257	Implantable cardiac pacemakers	E0601	Arrhythmia	A070101	Failure to Capture		
									A070909	Oversensing		
									A0712	Pacing Problem		
							F23	Unexpected Medical Intervention	A070101	Failure to Capture		
									A070909	Oversensing		
									A0712	Pacing Problem		
					CT2866	Implantable cardiac electrical stimulators	E0601	Arrhythmia	A070101	Failure to Capture		
									A070909	Oversensing		
									A0712	Pacing Problem		
							F23	Unexpected Medical Intervention	A070101	Failure to Capture		
									A070909	Oversensing		
									A0712	Pacing Problem		
16840	2025	All other reportable incidents	47265	Dual-chamber implantable pacemaker, rate-responsive			E2103	Device Overstimulation of Tissue	A0908	Incorrect, Inadequate or Imprecise Result or Readings	16/04/2025	16/04/2025
							F08	Hospitalization or Prolonged Hospitalization	A0908	Incorrect, Inadequate or Imprecise Result or Readings	16/04/2025	16/04/2025
							F15	Recognized Device or Procedural Complication	A0908	Incorrect, Inadequate or Imprecise Result or Readings	16/04/2025	16/04/2025
							F1905	Device Revision or Replacement	A0908	Incorrect, Inadequate or Imprecise Result or Readings	16/04/2025	16/04/2025

16841	2025	All other reportable incidents	47265	Dual-chamber implantable pacemaker, rate-responsive	CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	21/05/2025	21/05/2025
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	21/05/2025	21/05/2025
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	21/05/2025	21/05/2025
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	21/05/2025	21/05/2025
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	21/05/2025	21/05/2025
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	21/05/2025	21/05/2025
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	21/05/2025	21/05/2025
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	21/05/2025	21/05/2025
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	21/05/2025	21/05/2025
16842	2025	All other reportable incidents	47205	Cardiac pulse generator programmer	CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A150105	Key or Button Unresponsive/not Working		
							F26	No Health Consequences or Impact	A150105	Key or Button Unresponsive/not Working		
16843	2025	All other reportable incidents	60789	Intracardiac pacemaker	CT1685	Cardiac pacemakers	E0601	Arrhythmia	A24	No Apparent Device Problem		
							F12	Serious Injury/ Illness/ Impairment	A24	No Apparent Device Problem		
					CT257	Implantable cardiac pacemakers	E0601	Arrhythmia	A24	No Apparent Device Problem		
							F12	Serious Injury/ Illness/ Impairment	A24	No Apparent Device Problem		
					CT2818	Catheterization/Cannulation/Fluid path devices	E0601	Arrhythmia	A24	No Apparent Device Problem		
							F12	Serious Injury/ Illness/ Impairment	A24	No Apparent Device Problem		
					CT2866	Implantable cardiac electrical stimulators	E0601	Arrhythmia	A24	No Apparent Device Problem		
							F12	Serious Injury/ Illness/ Impairment	A24	No Apparent Device Problem		
16844	2025	All other reportable incidents	57968	Cardiovascular device introducer, steerable	CT1684	Cardiac pacemakers and associated devices	E0601	Arrhythmia	A24	No Apparent Device Problem		
							F12	Serious Injury/ Illness/ Impairment	A24	No Apparent Device Problem		
16845	2025	All other reportable incidents	35852	Single-chamber implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	23/04/2025	23/04/2025
							F1901	Additional Surgery	A070504	Premature Discharge of Battery	23/04/2025	23/04/2025
16846	2025	All other reportable incidents	35852	Single-chamber implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E210401	Shock from Patient Lead(s)	A070909	Oversensing	09/12/2024	09/12/2024
									A071302	Inappropriate/Inadequate Shock/Stimulation	09/12/2024	09/12/2024
									A090801	Signal Artifact/Noise	09/12/2024	09/12/2024
							F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A070909	Oversensing	09/12/2024	09/12/2024
									A071302	Inappropriate/Inadequate Shock/Stimulation	09/12/2024	09/12/2024
									A090801	Signal Artifact/Noise	09/12/2024	09/12/2024
16847	2025	All other reportable incidents	47265	Dual-chamber implantable pacemaker, rate-responsive	CT1685	Cardiac pacemakers	E060104	Bradycardia	A070909	Oversensing		
									A0712	Pacing Problem		
									A090801	Signal Artifact/Noise		
							F1001	Absence of Treatment	A070909	Oversensing		
									A0712	Pacing Problem		
									A090801	Signal Artifact/Noise		
					CT257	Implantable cardiac pacemakers	E060104	Bradycardia	A070909	Oversensing		
									A0712	Pacing Problem		
									A090801	Signal Artifact/Noise		
							F1001	Absence of Treatment	A070909	Oversensing		
									A0712	Pacing Problem		
									A090801	Signal Artifact/Noise		
					CT2866	Implantable cardiac electrical stimulators	E060104	Bradycardia	A070909	Oversensing		
									A0712	Pacing Problem		
									A090801	Signal Artifact/Noise		
							F1001	Absence of Treatment	A070909	Oversensing		
									A0712	Pacing Problem		
									A090801	Signal Artifact/Noise		
16848	2025	All other reportable incidents	47265	Dual-chamber implantable pacemaker, rate-responsive	CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem	22/04/2025	22/04/2025
									A070504	Premature Discharge of Battery	22/04/2025	22/04/2025
							F08	Hospitalization or Prolonged Hospitalization	A0705	Battery Problem	22/04/2025	22/04/2025
									A070504	Premature Discharge of Battery	22/04/2025	22/04/2025
							F1905	Device Revision or Replacement	A0705	Battery Problem	22/04/2025	22/04/2025
									A070504	Premature Discharge of Battery	22/04/2025	22/04/2025
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem	22/04/2025	22/04/2025
									A070504	Premature Discharge of Battery	22/04/2025	22/04/2025
							F08	Hospitalization or Prolonged Hospitalization	A0705	Battery Problem	22/04/2025	22/04/2025
									A070504	Premature Discharge of Battery	22/04/2025	22/04/2025
							F1905	Device Revision or Replacement	A0705	Battery Problem	22/04/2025	22/04/2025
									A070504	Premature Discharge of Battery	22/04/2025	22/04/2025
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem	22/04/2025	22/04/2025
									A070504	Premature Discharge of Battery	22/04/2025	22/04/2025
							F08	Hospitalization or Prolonged Hospitalization	A0705	Battery Problem	22/04/2025	22/04/2025
									A070504	Premature Discharge of Battery	22/04/2025	22/04/2025
							F1905	Device Revision or Replacement	A0705	Battery Problem	22/04/2025	22/04/2025
									A070504	Premature Discharge of Battery	22/04/2025	22/04/2025
16849	2025	All other reportable incidents	35852	Single-chamber implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A072201	High Impedance		
							F26	No Health Consequences or Impact	A072201	High Impedance		
16850	2025	All other reportable incidents	47270	Cardiac resynchronization therapy implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A160601	Premature Elective Replacement Indicator		
							F1905	Device Revision or Replacement	A160601	Premature Elective Replacement Indicator		
16851	2025	All other reportable incidents	17882	Physiologic-monitoring defibrillation system	CT1684	Cardiac pacemakers and associated devices	E2401	Insufficient Information	A2105	Inadequate or Insufficient Training		
							F05	Delay to Treatment/ Therapy	A2105	Inadequate or Insufficient Training		
16852	2025	All other reportable incidents	37265	Dual-chamber implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E210401	Shock from Patient Lead(s)	A0712	Pacing Problem		
									A071302	Inappropriate/Inadequate Shock/Stimulation		
							F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A0712	Pacing Problem		
									A071302	Inappropriate/Inadequate Shock/Stimulation		
16853	2025	All other reportable incidents	37265	Dual-chamber implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A070102	High Capture Threshold		
									A070909	Oversensing		
									A072202	Low Impedance		
									A090801	Signal Artifact/Noise		
									A1301	Failure to Read Input Signal		
							F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A070102	High Capture Threshold		
									A070909	Oversensing		
									A072202	Low Impedance		
									A090801	Signal Artifact/Noise		
									A1301	Failure to Read Input Signal		
16854	2025	All other reportable incidents	47270	Cardiac resynchronization therapy implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E210401	Shock from Patient Lead(s)	A070909	Oversensing		
									A071302	Inappropriate/Inadequate Shock/Stimulation		
							F08	Hospitalization or Prolonged Hospitalization	A070909	Oversensing		

									A071302	Inappropriate/Inadequate Shock/Stimulation		
							F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A070909	Oversensing		
									A071302	Inappropriate/Inadequate Shock/Stimulation		
16855	2025		47265	Dual-chamber implantable pacemaker, rate-responsive	CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	19/03/2025	19/03/2025
							F1901	Additional Surgery	A24	No Apparent Device Problem	19/03/2025	19/03/2025
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	19/03/2025	19/03/2025
							F1901	Additional Surgery	A24	No Apparent Device Problem	19/03/2025	19/03/2025
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	19/03/2025	19/03/2025
							F1901	Additional Surgery	A24	No Apparent Device Problem	19/03/2025	19/03/2025
16856	2025	All other reportable incidents	35852	Single-chamber implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A0508	Noise, Audible	17/04/2025	17/04/2025
									A070603	Failure to Charge	17/04/2025	17/04/2025
							F1901	Additional Surgery	A0508	Noise, Audible	17/04/2025	17/04/2025
									A070603	Failure to Charge	17/04/2025	17/04/2025
16857	2025	All other reportable incidents	47265	Dual-chamber implantable pacemaker, rate-responsive	CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
16858	2025	All other reportable incidents	47270	Cardiac resynchronization therapy implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A1301	Failure to Read Input Signal	01/04/2025	01/04/2025
							F08	Hospitalization or Prolonged Hospitalization	A1301	Failure to Read Input Signal	01/04/2025	01/04/2025
							F1001	Absence of Treatment	A1301	Failure to Read Input Signal	01/04/2025	01/04/2025
							F1906	Modified Surgical Procedure	A1301	Failure to Read Input Signal	01/04/2025	01/04/2025
							F1908	Prolonged Surgery	A1301	Failure to Read Input Signal	01/04/2025	01/04/2025
16859	2025	All other reportable incidents	47205	Cardiac pulse generator programmer	CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A0902	Display or Visual Feedback Problem		
							F2601	Problem Identified Before Clinical Use/Exposure	A0902	Display or Visual Feedback Problem		
16860	2025		17882	Physiologic-monitoring defibrillation system	CT1684	Cardiac pacemakers and associated devices	E2401	Insufficient Information	A25	No Apparent Adverse Event		
							F24	Insufficient Health Impact Information	A25	No Apparent Adverse Event		
16861	2025	All other reportable incidents	47267	Single-chamber implantable pacemaker, rate-responsive	CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	14/04/2025	14/04/2025
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	14/04/2025	14/04/2025
							F12	Serious Injury/ Illness/ Impairment	A070504	Premature Discharge of Battery	14/04/2025	14/04/2025
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	14/04/2025	14/04/2025
							F23	Unexpected Medical Intervention	A070504	Premature Discharge of Battery	14/04/2025	14/04/2025
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	14/04/2025	14/04/2025
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	14/04/2025	14/04/2025
							F12	Serious Injury/ Illness/ Impairment	A070504	Premature Discharge of Battery	14/04/2025	14/04/2025
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	14/04/2025	14/04/2025
							F23	Unexpected Medical Intervention	A070504	Premature Discharge of Battery	14/04/2025	14/04/2025
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	14/04/2025	14/04/2025
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	14/04/2025	14/04/2025
							F12	Serious Injury/ Illness/ Impairment	A070504	Premature Discharge of Battery	14/04/2025	14/04/2025
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	14/04/2025	14/04/2025
							F23	Unexpected Medical Intervention	A070504	Premature Discharge of Battery	14/04/2025	14/04/2025
16862	2025	All other reportable incidents	17882	Physiologic-monitoring defibrillation system	CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A0719	Unexpected Shutdown		
							F26	No Health Consequences or Impact	A0719	Unexpected Shutdown		
16863	2025		37265	Dual-chamber implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E0102	Brain Injury	A071301	Failure to Deliver Shock/Stimulation		
							F02	Death	A071301	Failure to Deliver Shock/Stimulation		
							F08	Hospitalization or Prolonged Hospitalization	A071301	Failure to Deliver Shock/Stimulation		
16864	2025	All other reportable incidents	47270	Cardiac resynchronization therapy implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A0709	Device Sensing Problem		
									A0712	Pacing Problem		
									A071302	Inappropriate/Inadequate Shock/Stimulation		
							F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A0709	Device Sensing Problem		
									A0712	Pacing Problem		
									A071302	Inappropriate/Inadequate Shock/Stimulation		
							F1203	Life Threatening Illness or Injury	A0709	Device Sensing Problem		
									A0712	Pacing Problem		
									A071302	Inappropriate/Inadequate Shock/Stimulation		
							F2307	Reprogramming of Device	A0709	Device Sensing Problem		
									A0712	Pacing Problem		
									A071302	Inappropriate/Inadequate Shock/Stimulation		
16865	2025	All other reportable incidents	47265	Dual-chamber implantable pacemaker, rate-responsive	CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A072001	Electromagnetic Interference		
							F26	No Health Consequences or Impact	A072001	Electromagnetic Interference		
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A072001	Electromagnetic Interference		
							F26	No Health Consequences or Impact	A072001	Electromagnetic Interference		
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A072001	Electromagnetic Interference		
							F26	No Health Consequences or Impact	A072001	Electromagnetic Interference		
16866	2025	All other reportable incidents	17882	Physiologic-monitoring defibrillation system	CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A0708	Power Problem		
							F27	Problem Identified During Non-Clinical Procedure	A0708	Power Problem		
16867	2025		47263	Cardiac resynchronization therapy implantable pacemaker	CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A070909	Oversensing		
									A070910	Undersensing		
									A071204	No Pacing		
									A071206	Pacing Asynchronously		
									A090801	Signal Artifact/Noise		
							F26	No Health Consequences or Impact	A070909	Oversensing		
									A070910	Undersensing		
									A071204	No Pacing		
									A071206	Pacing Asynchronously		
									A090801	Signal Artifact/Noise		
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A070909	Oversensing		
									A070910	Undersensing		
									A071204	No Pacing		
									A071206	Pacing Asynchronously		
									A090801	Signal Artifact/Noise		
							F26	No Health Consequences or Impact	A070909	Oversensing		
									A070910	Undersensing		
									A071204	No Pacing		
									A071206	Pacing Asynchronously		
									A090801	Signal Artifact/Noise		

					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A070909	Oversensing		
									A070910	Undersensing		
									A071204	No Pacing		
									A071206	Pacing Asynchronously		
									A090801	Signal Artifact/Noise		
							F26	No Health Consequences or Impact	A070909	Oversensing		
									A070910	Undersensing		
									A071204	No Pacing		
									A071206	Pacing Asynchronously		
									A090801	Signal Artifact/Noise		
16868	2025	All other reportable incidents	35852	Single-chamber implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E210401	Shock from Patient Lead(s)	A070909	Oversensing	25/05/2025	25/05/2025
									A071302	Inappropriate/Inadequate Shock/Stimulation	25/05/2025	25/05/2025
									A072001	Electromagnetic Interference	25/05/2025	25/05/2025
									A090801	Signal Artifact/Noise	25/05/2025	25/05/2025
							F08	Hospitalization or Prolonged Hospitalization	A070909	Oversensing	25/05/2025	25/05/2025
									A071302	Inappropriate/Inadequate Shock/Stimulation	25/05/2025	25/05/2025
									A072001	Electromagnetic Interference	25/05/2025	25/05/2025
									A090801	Signal Artifact/Noise	25/05/2025	25/05/2025
							F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A070909	Oversensing	25/05/2025	25/05/2025
									A071302	Inappropriate/Inadequate Shock/Stimulation	25/05/2025	25/05/2025
									A072001	Electromagnetic Interference	25/05/2025	25/05/2025
									A090801	Signal Artifact/Noise	25/05/2025	25/05/2025
							F1203	Life Threatening Illness or Injury	A070909	Oversensing	25/05/2025	25/05/2025
									A071302	Inappropriate/Inadequate Shock/Stimulation	25/05/2025	25/05/2025
									A072001	Electromagnetic Interference	25/05/2025	25/05/2025
									A090801	Signal Artifact/Noise	25/05/2025	25/05/2025
							F1903	Device Explantation	A070909	Oversensing	25/05/2025	25/05/2025
									A071302	Inappropriate/Inadequate Shock/Stimulation	25/05/2025	25/05/2025
									A072001	Electromagnetic Interference	25/05/2025	25/05/2025
									A090801	Signal Artifact/Noise	25/05/2025	25/05/2025
							F2203	Imaging Required	A070909	Oversensing	25/05/2025	25/05/2025
									A071302	Inappropriate/Inadequate Shock/Stimulation	25/05/2025	25/05/2025
									A072001	Electromagnetic Interference	25/05/2025	25/05/2025
									A090801	Signal Artifact/Noise	25/05/2025	25/05/2025
							F2307	Reprogramming of Device	A070909	Oversensing	25/05/2025	25/05/2025
									A071302	Inappropriate/Inadequate Shock/Stimulation	25/05/2025	25/05/2025
									A072001	Electromagnetic Interference	25/05/2025	25/05/2025
									A090801	Signal Artifact/Noise	25/05/2025	25/05/2025
16869	2025	All other reportable incidents	47270	Cardiac resynchronization therapy implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
									A0908	Incorrect, Inadequate or Imprecise Result or Readings		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
									A0908	Incorrect, Inadequate or Imprecise Result or Readings		
16870	2025	All other reportable incidents	37265	Dual-chamber implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A070910	Undersensing		
									A0713	Defibrillation/Stimulation Problem		
							F05	Delay to Treatment/ Therapy	A070910	Undersensing		
									A0713	Defibrillation/Stimulation Problem		
16871	2025	All other reportable incidents	47263	Cardiac resynchronization therapy implantable pacemaker	CT1685	Cardiac pacemakers	E0112	Dizziness	A070909	Oversensing	15/04/2025	15/04/2025
									A0712	Pacing Problem	15/04/2025	15/04/2025
									A0908	Incorrect, Inadequate or Imprecise Result or Readings	15/04/2025	15/04/2025
							E020204	Malaise	A070909	Oversensing	15/04/2025	15/04/2025
									A0712	Pacing Problem	15/04/2025	15/04/2025
									A0908	Incorrect, Inadequate or Imprecise Result or Readings	15/04/2025	15/04/2025
							E060101	Asystole	A070909	Oversensing	15/04/2025	15/04/2025
									A0712	Pacing Problem	15/04/2025	15/04/2025
									A0908	Incorrect, Inadequate or Imprecise Result or Readings	15/04/2025	15/04/2025
							F08	Hospitalization or Prolonged Hospitalization	A070909	Oversensing	15/04/2025	15/04/2025
									A0712	Pacing Problem	15/04/2025	15/04/2025
									A0908	Incorrect, Inadequate or Imprecise Result or Readings	15/04/2025	15/04/2025
							F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A070909	Oversensing	15/04/2025	15/04/2025
									A0712	Pacing Problem	15/04/2025	15/04/2025
									A0908	Incorrect, Inadequate or Imprecise Result or Readings	15/04/2025	15/04/2025
							F1001	Absence of Treatment	A070909	Oversensing	15/04/2025	15/04/2025
									A0712	Pacing Problem	15/04/2025	15/04/2025
									A0908	Incorrect, Inadequate or Imprecise Result or Readings	15/04/2025	15/04/2025
							F1905	Device Revision or Replacement	A070909	Oversensing	15/04/2025	15/04/2025
									A0712	Pacing Problem	15/04/2025	15/04/2025
									A0908	Incorrect, Inadequate or Imprecise Result or Readings	15/04/2025	15/04/2025
					CT257	Implantable cardiac pacemakers	E0112	Dizziness	A070909	Oversensing	15/04/2025	15/04/2025
									A0712	Pacing Problem	15/04/2025	15/04/2025
									A0908	Incorrect, Inadequate or Imprecise Result or Readings	15/04/2025	15/04/2025
							E020204	Malaise	A070909	Oversensing	15/04/2025	15/04/2025
									A0712	Pacing Problem	15/04/2025	15/04/2025
									A0908	Incorrect, Inadequate or Imprecise Result or Readings	15/04/2025	15/04/2025
							E060101	Asystole	A070909	Oversensing	15/04/2025	15/04/2025
									A0712	Pacing Problem	15/04/2025	15/04/2025
									A0908	Incorrect, Inadequate or Imprecise Result or Readings	15/04/2025	15/04/2025
							F08	Hospitalization or Prolonged Hospitalization	A070909	Oversensing	15/04/2025	15/04/2025
									A0712	Pacing Problem	15/04/2025	15/04/2025

									A0908	Incorrect, Inadequate or Imprecise Result or Readings	15/04/2025	15/04/2025
							F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A070909	Oversensing	15/04/2025	15/04/2025
									A0712	Pacing Problem	15/04/2025	15/04/2025
									A0908	Incorrect, Inadequate or Imprecise Result or Readings	15/04/2025	15/04/2025
							F1001	Absence of Treatment	A070909	Oversensing	15/04/2025	15/04/2025
									A0712	Pacing Problem	15/04/2025	15/04/2025
									A0908	Incorrect, Inadequate or Imprecise Result or Readings	15/04/2025	15/04/2025
							F1905	Device Revision or Replacement	A070909	Oversensing	15/04/2025	15/04/2025
									A0712	Pacing Problem	15/04/2025	15/04/2025
									A0908	Incorrect, Inadequate or Imprecise Result or Readings	15/04/2025	15/04/2025
					CT2866	Implantable cardiac electrical stimulators	E0112	Dizziness	A070909	Oversensing	15/04/2025	15/04/2025
									A0712	Pacing Problem	15/04/2025	15/04/2025
									A0908	Incorrect, Inadequate or Imprecise Result or Readings	15/04/2025	15/04/2025
							E020204	Malaise	A070909	Oversensing	15/04/2025	15/04/2025
									A0712	Pacing Problem	15/04/2025	15/04/2025
									A0908	Incorrect, Inadequate or Imprecise Result or Readings	15/04/2025	15/04/2025
							E060101	Asystole	A070909	Oversensing	15/04/2025	15/04/2025
									A0712	Pacing Problem	15/04/2025	15/04/2025
									A0908	Incorrect, Inadequate or Imprecise Result or Readings	15/04/2025	15/04/2025
							F08	Hospitalization or Prolonged Hospitalization	A070909	Oversensing	15/04/2025	15/04/2025
									A0712	Pacing Problem	15/04/2025	15/04/2025
									A0908	Incorrect, Inadequate or Imprecise Result or Readings	15/04/2025	15/04/2025
							F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A070909	Oversensing	15/04/2025	15/04/2025
									A0712	Pacing Problem	15/04/2025	15/04/2025
									A0908	Incorrect, Inadequate or Imprecise Result or Readings	15/04/2025	15/04/2025
							F1001	Absence of Treatment	A070909	Oversensing	15/04/2025	15/04/2025
									A0712	Pacing Problem	15/04/2025	15/04/2025
									A0908	Incorrect, Inadequate or Imprecise Result or Readings	15/04/2025	15/04/2025
							F1905	Device Revision or Replacement	A070909	Oversensing	15/04/2025	15/04/2025
									A0712	Pacing Problem	15/04/2025	15/04/2025
									A0908	Incorrect, Inadequate or Imprecise Result or Readings	15/04/2025	15/04/2025
16872	2025	All other reportable incidents	47270	Cardiac resynchronization therapy implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
16873	2025	All other reportable incidents	47270	Cardiac resynchronization therapy implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	24/03/2025	24/03/2025
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	24/03/2025	24/03/2025
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	24/03/2025	24/03/2025
16874	2025	All other reportable incidents	47263	Cardiac resynchronization therapy implantable pacemaker	CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A0712	Pacing Problem	09/04/2025	09/04/2025
									A0908	Incorrect, Inadequate or Imprecise Result or Readings	09/04/2025	09/04/2025
							F08	Hospitalization or Prolonged Hospitalization	A0712	Pacing Problem	09/04/2025	09/04/2025
									A0908	Incorrect, Inadequate or Imprecise Result or Readings	09/04/2025	09/04/2025
							F1001	Absence of Treatment	A0712	Pacing Problem	09/04/2025	09/04/2025
									A0908	Incorrect, Inadequate or Imprecise Result or Readings	09/04/2025	09/04/2025
							F1905	Device Revision or Replacement	A0712	Pacing Problem	09/04/2025	09/04/2025
									A0908	Incorrect, Inadequate or Imprecise Result or Readings	09/04/2025	09/04/2025
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A0712	Pacing Problem	09/04/2025	09/04/2025
									A0908	Incorrect, Inadequate or Imprecise Result or Readings	09/04/2025	09/04/2025
							F08	Hospitalization or Prolonged Hospitalization	A0712	Pacing Problem	09/04/2025	09/04/2025
									A0908	Incorrect, Inadequate or Imprecise Result or Readings	09/04/2025	09/04/2025
							F1001	Absence of Treatment	A0712	Pacing Problem	09/04/2025	09/04/2025
									A0908	Incorrect, Inadequate or Imprecise Result or Readings	09/04/2025	09/04/2025
							F1905	Device Revision or Replacement	A0712	Pacing Problem	09/04/2025	09/04/2025
									A0908	Incorrect, Inadequate or Imprecise Result or Readings	09/04/2025	09/04/2025
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A0712	Pacing Problem	09/04/2025	09/04/2025
									A0908	Incorrect, Inadequate or Imprecise Result or Readings	09/04/2025	09/04/2025
							F08	Hospitalization or Prolonged Hospitalization	A0712	Pacing Problem	09/04/2025	09/04/2025
									A0908	Incorrect, Inadequate or Imprecise Result or Readings	09/04/2025	09/04/2025
							F1001	Absence of Treatment	A0712	Pacing Problem	09/04/2025	09/04/2025
									A0908	Incorrect, Inadequate or Imprecise Result or Readings	09/04/2025	09/04/2025
							F1905	Device Revision or Replacement	A0712	Pacing Problem	09/04/2025	09/04/2025
									A0908	Incorrect, Inadequate or Imprecise Result or Readings	09/04/2025	09/04/2025
16875	2025	All other reportable incidents	47263	Cardiac resynchronization therapy implantable pacemaker	CT1685	Cardiac pacemakers	E060101	Asystole	A070101	Failure to Capture		
							F26	No Health Consequences or Impact	A070101	Failure to Capture		
					CT257	Implantable cardiac pacemakers	E060101	Asystole	A070101	Failure to Capture		
							F26	No Health Consequences or Impact	A070101	Failure to Capture		
					CT2866	Implantable cardiac electrical stimulators	E060101	Asystole	A070101	Failure to Capture		
							F26	No Health Consequences or Impact	A070101	Failure to Capture		
16876	2025	All other reportable incidents	47270	Cardiac resynchronization therapy implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A1301	Failure to Read Input Signal	01/04/2025	01/04/2025
							F08	Hospitalization or Prolonged Hospitalization	A1301	Failure to Read Input Signal	01/04/2025	01/04/2025
							F1001	Absence of Treatment	A1301	Failure to Read Input Signal	01/04/2025	01/04/2025
							F1906	Modified Surgical Procedure	A1301	Failure to Read Input Signal	01/04/2025	01/04/2025

							F1908	Prolonged Surgery	A1301	Failure to Read Input Signal	01/04/2025	01/04/2025
16877	2025	All other reportable incidents	47263	Cardiac resynchronization therapy implantable pacemaker	CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
16878	2025	All other reportable incidents	17882	Physiologic-monitoring defibrillation system	CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A090805	Incorrect Measurement		
							F27	Problem Identified During Non-Clinical Procedure	A090805	Incorrect Measurement		
16879	2025	All other reportable incidents	17882	Physiologic-monitoring defibrillation system	CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A07	Electrical /Electronic Property Problem		
							F27	Problem Identified During Non-Clinical Procedure	A07	Electrical /Electronic Property Problem		
16880	2025	All other reportable incidents	17882	Physiologic-monitoring defibrillation system	CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A0908	Incorrect, Inadequate or Imprecise Result or Readings		
							F27	Problem Identified During Non-Clinical Procedure	A0908	Incorrect, Inadequate or Imprecise Result or Readings		
16881	2025	All other reportable incidents	17882	Physiologic-monitoring defibrillation system	CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A0908	Incorrect, Inadequate or Imprecise Result or Readings		
							F27	Problem Identified During Non-Clinical Procedure	A0908	Incorrect, Inadequate or Imprecise Result or Readings		
16882	2025	All other reportable incidents	17882	Physiologic-monitoring defibrillation system	CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A0719	Unexpected Shutdown		
							F27	Problem Identified During Non-Clinical Procedure	A0719	Unexpected Shutdown		
16883	2025	All other reportable incidents	47270	Cardiac resynchronization therapy implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E210401	Shock from Patient Lead(s)	A071302	Inappropriate/Inadequate Shock/Stimulation	23/06/2025	23/06/2025
							F08	Hospitalization or Prolonged Hospitalization	A071302	Inappropriate/Inadequate Shock/Stimulation	23/06/2025	23/06/2025
							F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A071302	Inappropriate/Inadequate Shock/Stimulation	23/06/2025	23/06/2025
							F1905	Device Revision or Replacement	A071302	Inappropriate/Inadequate Shock/Stimulation	23/06/2025	23/06/2025
							F2203	Imaging Required	A071302	Inappropriate/Inadequate Shock/Stimulation	23/06/2025	23/06/2025
							F23	Unexpected Medical Intervention	A071302	Inappropriate/Inadequate Shock/Stimulation	23/06/2025	23/06/2025
16884	2025	All other reportable incidents	47270	Cardiac resynchronization therapy implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E210401	Shock from Patient Lead(s)	A071302	Inappropriate/Inadequate Shock/Stimulation	23/06/2025	23/06/2025
							F08	Hospitalization or Prolonged Hospitalization	A071302	Inappropriate/Inadequate Shock/Stimulation	23/06/2025	23/06/2025
							F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A071302	Inappropriate/Inadequate Shock/Stimulation	23/06/2025	23/06/2025
							F1905	Device Revision or Replacement	A071302	Inappropriate/Inadequate Shock/Stimulation	23/06/2025	23/06/2025
							F2203	Imaging Required	A071302	Inappropriate/Inadequate Shock/Stimulation	23/06/2025	23/06/2025
							F23	Unexpected Medical Intervention	A071302	Inappropriate/Inadequate Shock/Stimulation	23/06/2025	23/06/2025
16885	2025	All other reportable incidents	47270	Cardiac resynchronization therapy implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
									A0908	Incorrect, Inadequate or Imprecise Result or Readings		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
									A0908	Incorrect, Inadequate or Imprecise Result or Readings		
16886	2025		47265	Dual-chamber implantable pacemaker, rate-responsive			E0112	Dizziness	A070909	Oversensing		
									A071204	No Pacing		
							F1901	Additional Surgery	A070909	Oversensing		
									A071204	No Pacing		
16887	2025	All other reportable incidents	35852	Single-chamber implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E210401	Shock from Patient Lead(s)	A010402	Migration		
									A070909	Oversensing		
									A071302	Inappropriate/Inadequate Shock/Stimulation		
									A072201	High Impedance		
									A1301	Failure to Read Input Signal		
							F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A010402	Migration		
									A070909	Oversensing		
									A071302	Inappropriate/Inadequate Shock/Stimulation		
									A072201	High Impedance		
									A1301	Failure to Read Input Signal		
							F2203	Imaging Required	A010402	Migration		
									A070909	Oversensing		
									A071302	Inappropriate/Inadequate Shock/Stimulation		
									A072201	High Impedance		
									A1301	Failure to Read Input Signal		
16888	2025	All other reportable incidents	63188	Left-ventricular wireless pacing system sensor/transmitter	CT145	Cardiovascular devices	E0623	Unspecified Heart Problem	A25	No Apparent Adverse Event		
							F26	No Health Consequences or Impact	A25	No Apparent Adverse Event		
					CT1577	Wireless monitoring system transmitters	E0623	Unspecified Heart Problem	A25	No Apparent Adverse Event		
							F26	No Health Consequences or Impact	A25	No Apparent Adverse Event		
					CT1684	Cardiac pacemakers and associated devices	E0623	Unspecified Heart Problem	A25	No Apparent Adverse Event		
							F26	No Health Consequences or Impact	A25	No Apparent Adverse Event		
					CT2727	Active implant devices/ancillaries	E0623	Unspecified Heart Problem	A25	No Apparent Adverse Event		
							F26	No Health Consequences or Impact	A25	No Apparent Adverse Event		
					CT2750	Implantable cardiac electrical stimulators and associated devices	E0623	Unspecified Heart Problem	A25	No Apparent Adverse Event		
							F26	No Health Consequences or Impact	A25	No Apparent Adverse Event		
					CT2885	Physiological monitoring/measuring system transmitters/receivers/transceivers	E0623	Unspecified Heart Problem	A25	No Apparent Adverse Event		
							F26	No Health Consequences or Impact	A25	No Apparent Adverse Event		
16889	2025	All other reportable incidents	47265	Dual-chamber implantable pacemaker, rate-responsive	CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A160601	Premature Elective Replacement Indicator	04/10/2024	04/10/2024
							F1903	Device Explantation	A160601	Premature Elective Replacement Indicator	04/10/2024	04/10/2024
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A160601	Premature Elective Replacement Indicator	04/10/2024	04/10/2024
							F1903	Device Explantation	A160601	Premature Elective Replacement Indicator	04/10/2024	04/10/2024
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A160601	Premature Elective Replacement Indicator	04/10/2024	04/10/2024
							F1903	Device Explantation	A160601	Premature Elective Replacement Indicator	04/10/2024	04/10/2024
16890	2025	All other reportable incidents	47265	Dual-chamber implantable pacemaker, rate-responsive	CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A05	Mechanical Problem	04/04/2025	04/04/2025
							F26	No Health Consequences or Impact	A05	Mechanical Problem	04/04/2025	04/04/2025
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A05	Mechanical Problem	04/04/2025	04/04/2025
							F26	No Health Consequences or Impact	A05	Mechanical Problem	04/04/2025	04/04/2025
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A05	Mechanical Problem	04/04/2025	04/04/2025
							F26	No Health Consequences or Impact	A05	Mechanical Problem	04/04/2025	04/04/2025
16891	2025	All other reportable incidents	37265	Dual-chamber implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A070101	Failure to Capture		
							F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A070101	Failure to Capture		
							F15	Recognized Device or Procedural Complication	A070101	Failure to Capture		
16892	2025	All other reportable incidents			CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070101	Failure to Capture		
									A070910	Undersensing		
							F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A070101	Failure to Capture		
									A070910	Undersensing		

16893	2025	All other reportable incidents	47265	Dual-chamber implantable pacemaker, rate-responsive	CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A0908	Incorrect, Inadequate or Imprecise Result or Readings	05/04/2025	05/04/2025
							F08	Hospitalization or Prolonged Hospitalization	A0908	Incorrect, Inadequate or Imprecise Result or Readings	05/04/2025	05/04/2025
							F1905	Device Revision or Replacement	A0908	Incorrect, Inadequate or Imprecise Result or Readings	05/04/2025	05/04/2025
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A0908	Incorrect, Inadequate or Imprecise Result or Readings	05/04/2025	05/04/2025
							F08	Hospitalization or Prolonged Hospitalization	A0908	Incorrect, Inadequate or Imprecise Result or Readings	05/04/2025	05/04/2025
							F1905	Device Revision or Replacement	A0908	Incorrect, Inadequate or Imprecise Result or Readings	05/04/2025	05/04/2025
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A0908	Incorrect, Inadequate or Imprecise Result or Readings	05/04/2025	05/04/2025
							F08	Hospitalization or Prolonged Hospitalization	A0908	Incorrect, Inadequate or Imprecise Result or Readings	05/04/2025	05/04/2025
							F1905	Device Revision or Replacement	A0908	Incorrect, Inadequate or Imprecise Result or Readings	05/04/2025	05/04/2025
16894	2025	All other reportable incidents	47270	Cardiac resynchronization therapy implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E210401	Shock from Patient Lead(s)	A010402	Migration		
									A070909	Oversensing		
									A071302	Inappropriate/Inadequate Shock/Stimulation		
									A072201	High Impedance		
									A1301	Failure to Read Input Signal		
							F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A010402	Migration		
									A070909	Oversensing		
									A071302	Inappropriate/Inadequate Shock/Stimulation		
									A072201	High Impedance		
									A1301	Failure to Read Input Signal		
							F2203	Imaging Required	A010402	Migration		
									A070909	Oversensing		
									A071302	Inappropriate/Inadequate Shock/Stimulation		
									A072201	High Impedance		
									A1301	Failure to Read Input Signal		
16895	2025	All other reportable incidents	47265	Dual-chamber implantable pacemaker, rate-responsive	CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	24/04/2025	24/04/2025
							F1901	Additional Surgery	A070504	Premature Discharge of Battery	24/04/2025	24/04/2025
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	24/04/2025	24/04/2025
							F1901	Additional Surgery	A070504	Premature Discharge of Battery	24/04/2025	24/04/2025
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	24/04/2025	24/04/2025
							F1901	Additional Surgery	A070504	Premature Discharge of Battery	24/04/2025	24/04/2025
16896	2025	All other reportable incidents	47265	Dual-chamber implantable pacemaker, rate-responsive	CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A070101	Failure to Capture		
									A070909	Oversensing		
							F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A070101	Failure to Capture		
									A070909	Oversensing		
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A070101	Failure to Capture		
									A070909	Oversensing		
							F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A070101	Failure to Capture		
									A070909	Oversensing		
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A070101	Failure to Capture		
									A070909	Oversensing		
							F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A070101	Failure to Capture		
									A070909	Oversensing		
16897	2025	All other reportable incidents	47267	Single-chamber implantable pacemaker, rate-responsive	CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A0104	Migration or Expulsion of Device		
							F08	Hospitalization or Prolonged Hospitalization	A0104	Migration or Expulsion of Device		
							F12	Serious Injury/ Illness/ Impairment	A0104	Migration or Expulsion of Device		
							F1905	Device Revision or Replacement	A0104	Migration or Expulsion of Device		
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A0104	Migration or Expulsion of Device		
							F08	Hospitalization or Prolonged Hospitalization	A0104	Migration or Expulsion of Device		
							F12	Serious Injury/ Illness/ Impairment	A0104	Migration or Expulsion of Device		
							F1905	Device Revision or Replacement	A0104	Migration or Expulsion of Device		
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A0104	Migration or Expulsion of Device		
							F08	Hospitalization or Prolonged Hospitalization	A0104	Migration or Expulsion of Device		
							F12	Serious Injury/ Illness/ Impairment	A0104	Migration or Expulsion of Device		
							F1905	Device Revision or Replacement	A0104	Migration or Expulsion of Device		
16898	2025	All other reportable incidents	36106	Implantable pacemaker analyser, intraoperative	CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A11	Computer Software Problem		
									A2204	Inadequate User Interface		
							F26	No Health Consequences or Impact	A11	Computer Software Problem		
									A2204	Inadequate User Interface		
16899	2025	All other reportable incidents	47265	Dual-chamber implantable pacemaker, rate-responsive	CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A05	Mechanical Problem	23/04/2025	23/04/2025
							F26	No Health Consequences or Impact	A05	Mechanical Problem	23/04/2025	23/04/2025
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A05	Mechanical Problem	23/04/2025	23/04/2025
							F26	No Health Consequences or Impact	A05	Mechanical Problem	23/04/2025	23/04/2025
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A05	Mechanical Problem	23/04/2025	23/04/2025
							F26	No Health Consequences or Impact	A05	Mechanical Problem	23/04/2025	23/04/2025
16900	2025	All other reportable incidents	47265	Dual-chamber implantable pacemaker, rate-responsive	CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A1605	Reset Problem		
							F12	Serious Injury/ Illness/ Impairment	A1605	Reset Problem		
							F2307	Reprogramming of Device	A1605	Reset Problem		
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A1605	Reset Problem		
							F12	Serious Injury/ Illness/ Impairment	A1605	Reset Problem		
							F2307	Reprogramming of Device	A1605	Reset Problem		
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A1605	Reset Problem		
							F12	Serious Injury/ Illness/ Impairment	A1605	Reset Problem		
							F2307	Reprogramming of Device	A1605	Reset Problem		
16901	2025	All other reportable incidents	47270	Cardiac resynchronization therapy implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A070909	Oversensing		
							F05	Delay to Treatment/ Therapy	A070909	Oversensing		
16902	2025	All other reportable incidents	47270	Cardiac resynchronization therapy implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E0601	Arrhythmia	A071202	Inaccurate Synchronization		
							F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A071202	Inaccurate Synchronization		
16903	2025	All other reportable incidents	47205	Cardiac pulse generator programmer	CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A07	Electrical /Electronic Property Problem		
							F26	No Health Consequences or Impact	A07	Electrical /Electronic Property Problem		
16904	2025	All other reportable incidents	47270	Cardiac resynchronization therapy implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E060110	Ventricular Fibrillation	A070909	Oversensing		
									A071202	Inaccurate Synchronization		

							F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A070909	Oversensing		
									A071202	Inaccurate Synchronization		
							F1203	Life Threatening Illness or Injury	A070909	Oversensing		
									A071202	Inaccurate Synchronization		
16905	2025	All other reportable incidents	47270	Cardiac resynchronization therapy implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E210401	Shock from Patient Lead(s)	A070909	Oversensing		
									A071302	Inappropriate/Inadequate Shock/Stimulation		
									A1301	Failure to Read Input Signal		
							E233001	Chest Pain	A070909	Oversensing		
									A071302	Inappropriate/Inadequate Shock/Stimulation		
									A1301	Failure to Read Input Signal		
							F08	Hospitalization or Prolonged Hospitalization	A070909	Oversensing		
									A071302	Inappropriate/Inadequate Shock/Stimulation		
									A1301	Failure to Read Input Signal		
							F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A070909	Oversensing		
									A071302	Inappropriate/Inadequate Shock/Stimulation		
									A1301	Failure to Read Input Signal		
16906	2025	All other reportable incidents	47270	Cardiac resynchronization therapy implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E210401	Shock from Patient Lead(s)	A070909	Oversensing		
									A071302	Inappropriate/Inadequate Shock/Stimulation		
									A090801	Signal Artifact/Noise		
							E233001	Chest Pain	A070909	Oversensing		
									A071302	Inappropriate/Inadequate Shock/Stimulation		
									A090801	Signal Artifact/Noise		
							F08	Hospitalization or Prolonged Hospitalization	A070909	Oversensing		
									A071302	Inappropriate/Inadequate Shock/Stimulation		
									A090801	Signal Artifact/Noise		
							F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A070909	Oversensing		
									A071302	Inappropriate/Inadequate Shock/Stimulation		
									A090801	Signal Artifact/Noise		
16907	2025	All other reportable incidents	47263	Cardiac resynchronization therapy implantable pacemaker	CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A160601	Premature Elective Replacement Indicator		
							F1903	Device Explantation	A160601	Premature Elective Replacement Indicator		
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A160601	Premature Elective Replacement Indicator		
							F1903	Device Explantation	A160601	Premature Elective Replacement Indicator		
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A160601	Premature Elective Replacement Indicator		
							F1903	Device Explantation	A160601	Premature Elective Replacement Indicator		
16908	2025	All other reportable incidents	17882	Physiologic-monitoring defibrillation system	CT1684	Cardiac pacemakers and associated devices	E0602	Cardiac Arrest	A0709	Device Sensing Problem		
							F26	No Health Consequences or Impact	A0709	Device Sensing Problem		
16909	2025	All other reportable incidents	17882	Physiologic-monitoring defibrillation system	CT1684	Cardiac pacemakers and associated devices	E0602	Cardiac Arrest	A0709	Device Sensing Problem		
							F26	No Health Consequences or Impact	A0709	Device Sensing Problem		
16910	2025	All other reportable incidents	47265	Dual-chamber implantable pacemaker, rate-responsive	CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	01/04/2025	01/04/2025
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	01/04/2025	01/04/2025
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	01/04/2025	01/04/2025
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	01/04/2025	01/04/2025
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	01/04/2025	01/04/2025
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	01/04/2025	01/04/2025
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	01/04/2025	01/04/2025
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	01/04/2025	01/04/2025
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	01/04/2025	01/04/2025
16911	2025	All other reportable incidents	17882	Physiologic-monitoring defibrillation system	CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A160503	Inappropriate or Unexpected Reset		
							F27	Problem Identified During Non-Clinical Procedure	A160503	Inappropriate or Unexpected Reset		
16912	2025	All other reportable incidents	47270	Cardiac resynchronization therapy implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E2104	Electric Shock	A0712	Pacing Problem		
									A071302	Inappropriate/Inadequate Shock/Stimulation		
							F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A0712	Pacing Problem		
									A071302	Inappropriate/Inadequate Shock/Stimulation		
16913	2025	All other reportable incidents	47270	Cardiac resynchronization therapy implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	25/01/2025	25/01/2025
									A0908	Incorrect, Inadequate or Imprecise Result or Readings	25/01/2025	25/01/2025
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	25/01/2025	25/01/2025
									A0908	Incorrect, Inadequate or Imprecise Result or Readings	25/01/2025	25/01/2025
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	25/01/2025	25/01/2025
									A0908	Incorrect, Inadequate or Imprecise Result or Readings	25/01/2025	25/01/2025
16914	2025	All other reportable incidents	47265	Dual-chamber implantable pacemaker, rate-responsive			E2103	Device Overstimulation of Tissue	A0712	Pacing Problem	02/04/2025	02/04/2025
									A0908	Incorrect, Inadequate or Imprecise Result or Readings	02/04/2025	02/04/2025
							E2311	Discomfort	A0712	Pacing Problem	02/04/2025	02/04/2025
									A0908	Incorrect, Inadequate or Imprecise Result or Readings	02/04/2025	02/04/2025
							F08	Hospitalization or Prolonged Hospitalization	A0712	Pacing Problem	02/04/2025	02/04/2025
									A0908	Incorrect, Inadequate or Imprecise Result or Readings	02/04/2025	02/04/2025
							F1905	Device Revision or Replacement	A0712	Pacing Problem	02/04/2025	02/04/2025
									A0908	Incorrect, Inadequate or Imprecise Result or Readings	02/04/2025	02/04/2025
16915	2025	All other reportable incidents	17882	Physiologic-monitoring defibrillation system	CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A0719	Unexpected Shutdown		
							F26	No Health Consequences or Impact	A0719	Unexpected Shutdown		
16916	2025	All other reportable incidents	47265	Dual-chamber implantable pacemaker, rate-responsive	CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A05	Mechanical Problem	24/04/2025	24/04/2025
							F26	No Health Consequences or Impact	A05	Mechanical Problem	24/04/2025	24/04/2025
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A05	Mechanical Problem	24/04/2025	24/04/2025
							F26	No Health Consequences or Impact	A05	Mechanical Problem	24/04/2025	24/04/2025
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A05	Mechanical Problem	24/04/2025	24/04/2025
							F26	No Health Consequences or Impact	A05	Mechanical Problem	24/04/2025	24/04/2025
16917	2025	All other reportable incidents	47270	Cardiac resynchronization therapy implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E011903	Syncope/Fainting	A070912	Incorrect Interpretation of Signal		
									A071301	Failure to Deliver Shock/Stimulation		
							F12	Serious Injury/ Illness/ Impairment	A070912	Incorrect Interpretation of Signal		
									A071301	Failure to Deliver Shock/Stimulation		
16918	2025	All other reportable incidents	37265	Dual-chamber implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E060109	Tachycardia	A0709	Device Sensing Problem		
									A071301	Failure to Deliver Shock/Stimulation		
							F1001	Absence of Treatment	A0709	Device Sensing Problem		

									A071301	Failure to Deliver Shock/Stimulation		
							F1203	Life Threatening Illness or Injury	A0709	Device Sensing Problem		
									A071301	Failure to Deliver Shock/Stimulation		
							F2307	Reprogramming of Device	A0709	Device Sensing Problem		
									A071301	Failure to Deliver Shock/Stimulation		
16919	2025	All other reportable incidents	35852	Single-chamber implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E210401	Shock from Patient Lead(s)	A071302	Inappropriate/Inadequate Shock/Stimulation		
							F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A071302	Inappropriate/Inadequate Shock/Stimulation		
16920	2025	All other reportable incidents	47265	Dual-chamber implantable pacemaker, rate-responsive	CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
16921	2025	All other reportable incidents	35852	Single-chamber implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A070101	Failure to Capture		
							F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A070101	Failure to Capture		
16922	2025	All other reportable incidents	47270	Cardiac resynchronization therapy implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E210401	Shock from Patient Lead(s)	A071302	Inappropriate/Inadequate Shock/Stimulation		
							F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A071302	Inappropriate/Inadequate Shock/Stimulation		
16923	2025	All other reportable incidents	37265	Dual-chamber implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem		
							F12	Serious Injury/ Illness/ Impairment	A0705	Battery Problem		
							F2307	Reprogramming of Device	A0705	Battery Problem		
16924	2025	All other reportable incidents	35852	Single-chamber implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
									A072201	High Impedance		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
									A072201	High Impedance		
16925	2025	All other reportable incidents	37265	Dual-chamber implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A070909	Oversensing		
									A0712	Pacing Problem		
							F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A070909	Oversensing		
									A0712	Pacing Problem		
16926	2025	All other reportable incidents	35852	Single-chamber implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A070909	Oversensing		
									A072001	Electromagnetic Interference		
									A090801	Signal Artifact/Noise		
							F12	Serious Injury/ Illness/ Impairment	A070909	Oversensing		
									A072001	Electromagnetic Interference		
									A090801	Signal Artifact/Noise		
							F2307	Reprogramming of Device	A070909	Oversensing		
									A072001	Electromagnetic Interference		
									A090801	Signal Artifact/Noise		
16927	2025	All other reportable incidents	47270	Cardiac resynchronization therapy implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A0709	Device Sensing Problem		
									A070909	Oversensing		
									A070910	Undersensing		
									A090801	Signal Artifact/Noise		
									A12	Connection Problem		
									A150202	Malposition of Device		
							F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A0709	Device Sensing Problem		
									A070909	Oversensing		
									A070910	Undersensing		
									A090801	Signal Artifact/Noise		
									A12	Connection Problem		
									A150202	Malposition of Device		
							F2203	Imaging Required	A0709	Device Sensing Problem		
									A070909	Oversensing		
									A070910	Undersensing		
									A090801	Signal Artifact/Noise		
									A12	Connection Problem		
									A150202	Malposition of Device		
16928	2025	All other reportable incidents	12913	Pacemaker, implantable	CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A070102	High Capture Threshold	01/05/2025	01/05/2025
									A070504	Premature Discharge of Battery	01/05/2025	01/05/2025
							F08	Hospitalization or Prolonged Hospitalization	A070102	High Capture Threshold	01/05/2025	01/05/2025
									A070504	Premature Discharge of Battery	01/05/2025	01/05/2025
							F1905	Device Revision or Replacement	A070102	High Capture Threshold	01/05/2025	01/05/2025
									A070504	Premature Discharge of Battery	01/05/2025	01/05/2025
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A070102	High Capture Threshold	01/05/2025	01/05/2025
									A070504	Premature Discharge of Battery	01/05/2025	01/05/2025
							F08	Hospitalization or Prolonged Hospitalization	A070102	High Capture Threshold	01/05/2025	01/05/2025
									A070504	Premature Discharge of Battery	01/05/2025	01/05/2025
							F1905	Device Revision or Replacement	A070102	High Capture Threshold	01/05/2025	01/05/2025
									A070504	Premature Discharge of Battery	01/05/2025	01/05/2025
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A070102	High Capture Threshold	01/05/2025	01/05/2025
									A070504	Premature Discharge of Battery	01/05/2025	01/05/2025
							F08	Hospitalization or Prolonged Hospitalization	A070102	High Capture Threshold	01/05/2025	01/05/2025
									A070504	Premature Discharge of Battery	01/05/2025	01/05/2025
							F1905	Device Revision or Replacement	A070102	High Capture Threshold	01/05/2025	01/05/2025
									A070504	Premature Discharge of Battery	01/05/2025	01/05/2025
16929	2025	All other reportable incidents	47270	Cardiac resynchronization therapy implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E210401	Shock from Patient Lead(s)	A070909	Oversensing		
									A071302	Inappropriate/Inadequate Shock/Stimulation		
									A090801	Signal Artifact/Noise		
							F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A070909	Oversensing		
									A071302	Inappropriate/Inadequate Shock/Stimulation		
									A090801	Signal Artifact/Noise		
16930	2025	All other reportable incidents	47270	Cardiac resynchronization therapy implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	16/05/2025	16/05/2025
									A0908	Incorrect, Inadequate or Imprecise Result or Readings	16/05/2025	16/05/2025
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	16/05/2025	16/05/2025
									A0908	Incorrect, Inadequate or Imprecise Result or Readings	16/05/2025	16/05/2025
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	16/05/2025	16/05/2025
									A0908	Incorrect, Inadequate or Imprecise Result or Readings	16/05/2025	16/05/2025

16931	2025	All other reportable incidents	37265	Dual-chamber implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A160601	Premature Elective Replacement Indicator		
							F1905	Device Revision or Replacement	A160601	Premature Elective Replacement Indicator		
16932	2025	All other reportable incidents	17882	Physiologic-monitoring defibrillation system	CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A0719	Unexpected Shutdown		
							F27	Problem Identified During Non-Clinical Procedure	A0719	Unexpected Shutdown		
16933	2025	All other reportable incidents	17882	Physiologic-monitoring defibrillation system	CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A0719	Unexpected Shutdown		
							F27	Problem Identified During Non-Clinical Procedure	A0719	Unexpected Shutdown		
16934	2025	All other reportable incidents	47265	Dual-chamber implantable pacemaker, rate-responsive	CT1685	Cardiac pacemakers	E2311	Discomfort	A070909	Oversensing		
									A0712	Pacing Problem		
									A12	Connection Problem		
							F1901	Additional Surgery	A070909	Oversensing		
									A0712	Pacing Problem		
									A12	Connection Problem		
					CT257	Implantable cardiac pacemakers	E2311	Discomfort	A070909	Oversensing		
									A0712	Pacing Problem		
									A12	Connection Problem		
							F1901	Additional Surgery	A070909	Oversensing		
									A0712	Pacing Problem		
									A12	Connection Problem		
					CT2866	Implantable cardiac electrical stimulators	E2311	Discomfort	A070909	Oversensing		
									A0712	Pacing Problem		
									A12	Connection Problem		
							F1901	Additional Surgery	A070909	Oversensing		
									A0712	Pacing Problem		
									A12	Connection Problem		
16935	2025	All other reportable incidents	47270	Cardiac resynchronization therapy implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A070909	Oversensing		
							F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A070909	Oversensing		
16936	2025	All other reportable incidents	47270	Cardiac resynchronization therapy implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E2104	Electric Shock	A071201	Failure to Convert Rhythm		
							F08	Hospitalization or Prolonged Hospitalization	A071201	Failure to Convert Rhythm		
							F1905	Device Revision or Replacement	A071201	Failure to Convert Rhythm		
							F1908	Prolonged Surgery	A071201	Failure to Convert Rhythm		
							F22	Unexpected Diagnostic Intervention	A071201	Failure to Convert Rhythm		
							F2203	Imaging Required	A071201	Failure to Convert Rhythm		
							F2306	Resuscitation	A071201	Failure to Convert Rhythm		
16937	2025	All other reportable incidents	47270	Cardiac resynchronization therapy implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E2104	Electric Shock	A0101	Patient-Device Incompatibility		
									A071201	Failure to Convert Rhythm		
							F08	Hospitalization or Prolonged Hospitalization	A0101	Patient-Device Incompatibility		
									A071201	Failure to Convert Rhythm		
							F1905	Device Revision or Replacement	A0101	Patient-Device Incompatibility		
									A071201	Failure to Convert Rhythm		
							F1908	Prolonged Surgery	A0101	Patient-Device Incompatibility		
									A071201	Failure to Convert Rhythm		
							F22	Unexpected Diagnostic Intervention	A0101	Patient-Device Incompatibility		
									A071201	Failure to Convert Rhythm		
							F2203	Imaging Required	A0101	Patient-Device Incompatibility		
									A071201	Failure to Convert Rhythm		
							F2306	Resuscitation	A0101	Patient-Device Incompatibility		
									A071201	Failure to Convert Rhythm		
16938	2025	All other reportable incidents	47270	Cardiac resynchronization therapy implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A070909	Oversensing		
									A090801	Signal Artifact/Noise		
							F23	Unexpected Medical Intervention	A070909	Oversensing		
									A090801	Signal Artifact/Noise		
16939	2025	All other reportable incidents	47270	Cardiac resynchronization therapy implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A070909	Oversensing		
									A090801	Signal Artifact/Noise		
							F23	Unexpected Medical Intervention	A070909	Oversensing		
									A090801	Signal Artifact/Noise		
16940	2025		47265	Dual-chamber implantable pacemaker, rate-responsive	CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem		
							F1901	Additional Surgery	A24	No Apparent Device Problem		
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem		
							F1901	Additional Surgery	A24	No Apparent Device Problem		
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem		
							F1901	Additional Surgery	A24	No Apparent Device Problem		
16941	2025	All other reportable incidents	47265	Dual-chamber implantable pacemaker, rate-responsive	CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A071102	Failure to Interrogate		
							F26	No Health Consequences or Impact	A071102	Failure to Interrogate		
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A071102	Failure to Interrogate		
							F26	No Health Consequences or Impact	A071102	Failure to Interrogate		
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A071102	Failure to Interrogate		
							F26	No Health Consequences or Impact	A071102	Failure to Interrogate		
16942	2025	All other reportable incidents	36106	Implantable pacemaker analyser, intraoperative	CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A0902	Display or Visual Feedback Problem		
									A11	Computer Software Problem		
									A13	Communication or Transmission Problem		
							F26	No Health Consequences or Impact	A0902	Display or Visual Feedback Problem		
									A11	Computer Software Problem		
									A13	Communication or Transmission Problem		
16943	2025	All other reportable incidents	47265	Dual-chamber implantable pacemaker, rate-responsive	CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A0701	Capturing Problem		
									A0705	Battery Problem		
									A070909	Oversensing		
							F26	No Health Consequences or Impact	A0701	Capturing Problem		
									A0705	Battery Problem		
									A070909	Oversensing		
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A0701	Capturing Problem		
									A0705	Battery Problem		
									A070909	Oversensing		
							F26	No Health Consequences or Impact	A0701	Capturing Problem		
									A0705	Battery Problem		
									A070909	Oversensing		
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A0701	Capturing Problem		
									A0705	Battery Problem		
									A070909	Oversensing		

							F26	No Health Consequences or Impact	A0701	Capturing Problem		
									A0705	Battery Problem		
									A070909	Oversensing		
16944	2025	All other reportable incidents	47265	Dual-chamber implantable pacemaker, rate-responsive	CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
16945	2025	All other reportable incidents	47270	Cardiac resynchronization therapy implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E060103	Atrial Flutter	A0709	Device Sensing Problem		
									A070909	Oversensing		
									A090801	Signal Artifact/Noise		
							F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A0709	Device Sensing Problem		
									A070909	Oversensing		
									A090801	Signal Artifact/Noise		
16946	2025	All other reportable incidents	47265	Dual-chamber implantable pacemaker, rate-responsive	CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	26/02/2025	26/02/2025
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	26/02/2025	26/02/2025
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	26/02/2025	26/02/2025
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	26/02/2025	26/02/2025
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	26/02/2025	26/02/2025
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	26/02/2025	26/02/2025
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	26/02/2025	26/02/2025
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	26/02/2025	26/02/2025
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	26/02/2025	26/02/2025
16947	2025	All other reportable incidents	47265	Dual-chamber implantable pacemaker, rate-responsive	CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
16948	2025	All other reportable incidents	47270	Cardiac resynchronization therapy implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E0601	Arrhythmia	A0712	Pacing Problem		
							E060109	Tachycardia	A0712	Pacing Problem		
							F12	Serious Injury/ Illness/ Impairment	A0712	Pacing Problem		
							F2307	Reprogramming of Device	A0712	Pacing Problem		
16949	2025	All other reportable incidents	47205	Cardiac pulse generator programmer	CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A13	Communication or Transmission Problem		
							F26	No Health Consequences or Impact	A13	Communication or Transmission Problem		
16950	2025	All other reportable incidents	47205	Cardiac pulse generator programmer	CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A13	Communication or Transmission Problem		
							F26	No Health Consequences or Impact	A13	Communication or Transmission Problem		
16951	2025	All other reportable incidents	47270	Cardiac resynchronization therapy implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	28/05/2025	28/05/2025
									A0908	Incorrect, Inadequate or Imprecise Result or Readings	28/05/2025	28/05/2025
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery	28/05/2025	28/05/2025
									A0908	Incorrect, Inadequate or Imprecise Result or Readings	28/05/2025	28/05/2025
16952	2025	All other reportable incidents	37265	Dual-chamber implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E060109	Tachycardia	A0709	Device Sensing Problem		
									A0712	Pacing Problem		
									A071301	Failure to Deliver Shock/Stimulation		
							F1001	Absence of Treatment	A0709	Device Sensing Problem		
									A0712	Pacing Problem		
									A071301	Failure to Deliver Shock/Stimulation		
							F1203	Life Threatening Illness or Injury	A0709	Device Sensing Problem		
									A0712	Pacing Problem		
									A071301	Failure to Deliver Shock/Stimulation		
							F2307	Reprogramming of Device	A0709	Device Sensing Problem		
									A0712	Pacing Problem		
									A071301	Failure to Deliver Shock/Stimulation		
16953	2025	All other reportable incidents	60789	Intracardiac pacemaker	CT1685	Cardiac pacemakers	E011903	Syncope/Fainting	A051201	Device Dislodged or Dislocated		
									A070101	Failure to Capture		
									A070908	Failure to Sense		
							E0503	Embolism/Embolus	A051201	Device Dislodged or Dislocated		
									A070101	Failure to Capture		
									A070908	Failure to Sense		
							F1901	Additional Surgery	A051201	Device Dislodged or Dislocated		
									A070101	Failure to Capture		
									A070908	Failure to Sense		
					CT257	Implantable cardiac pacemakers	E011903	Syncope/Fainting	A051201	Device Dislodged or Dislocated		
									A070101	Failure to Capture		
									A070908	Failure to Sense		
							E0503	Embolism/Embolus	A051201	Device Dislodged or Dislocated		
									A070101	Failure to Capture		
									A070908	Failure to Sense		
							F1901	Additional Surgery	A051201	Device Dislodged or Dislocated		
									A070101	Failure to Capture		
									A070908	Failure to Sense		
					CT2818	Catheterization/Cannulation/Fluid path devices	E011903	Syncope/Fainting	A051201	Device Dislodged or Dislocated		
									A070101	Failure to Capture		
									A070908	Failure to Sense		
							E0503	Embolism/Embolus	A051201	Device Dislodged or Dislocated		
									A070101	Failure to Capture		
									A070908	Failure to Sense		
							F1901	Additional Surgery	A051201	Device Dislodged or Dislocated		
									A070101	Failure to Capture		

									A070908	Failure to Sense		
					CT2866	Implantable cardiac electrical stimulators	E011903	Syncope/Fainting	A051201	Device Dislodged or Dislocated		
									A070101	Failure to Capture		
									A070908	Failure to Sense		
							E0503	Embolism/Embolus	A051201	Device Dislodged or Dislocated		
									A070101	Failure to Capture		
									A070908	Failure to Sense		
							F1901	Additional Surgery	A051201	Device Dislodged or Dislocated		
									A070101	Failure to Capture		
									A070908	Failure to Sense		
16954	2025	All other reportable incidents	47270	Cardiac resynchronization therapy implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	27/05/2025	27/05/2025
									A0908	Incorrect, Inadequate or Imprecise Result or Readings	27/05/2025	27/05/2025
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	27/05/2025	27/05/2025
									A0908	Incorrect, Inadequate or Imprecise Result or Readings	27/05/2025	27/05/2025
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	27/05/2025	27/05/2025
									A0908	Incorrect, Inadequate or Imprecise Result or Readings	27/05/2025	27/05/2025
16955	2025	All other reportable incidents	47270	Cardiac resynchronization therapy implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A0709	Device Sensing Problem		
									A070909	Oversensing		
									A090801	Signal Artifact/Noise		
							F26	No Health Consequences or Impact	A0709	Device Sensing Problem		
									A070909	Oversensing		
									A090801	Signal Artifact/Noise		
16956	2025	All other reportable incidents	47265	Dual-chamber implantable pacemaker, rate-responsive	CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A070910	Undersensing		
									A1301	Failure to Read Input Signal		
							F05	Delay to Treatment/ Therapy	A070910	Undersensing		
									A1301	Failure to Read Input Signal		
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A070910	Undersensing		
									A1301	Failure to Read Input Signal		
							F05	Delay to Treatment/ Therapy	A070910	Undersensing		
									A1301	Failure to Read Input Signal		
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A070910	Undersensing		
									A1301	Failure to Read Input Signal		
							F05	Delay to Treatment/ Therapy	A070910	Undersensing		
									A1301	Failure to Read Input Signal		
16957	2025	All other reportable incidents	35852	Single-chamber implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem	18/06/2025	18/06/2025
									A070504	Premature Discharge of Battery	18/06/2025	18/06/2025
									A0908	Incorrect, Inadequate or Imprecise Result or Readings	18/06/2025	18/06/2025
							F08	Hospitalization or Prolonged Hospitalization	A0705	Battery Problem	18/06/2025	18/06/2025
									A070504	Premature Discharge of Battery	18/06/2025	18/06/2025
									A0908	Incorrect, Inadequate or Imprecise Result or Readings	18/06/2025	18/06/2025
							F1905	Device Revision or Replacement	A0705	Battery Problem	18/06/2025	18/06/2025
									A070504	Premature Discharge of Battery	18/06/2025	18/06/2025
									A0908	Incorrect, Inadequate or Imprecise Result or Readings	18/06/2025	18/06/2025
16958	2025	All other reportable incidents	17882	Physiologic-monitoring defibrillation system	CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A090805	Incorrect Measurement		
							F27	Problem Identified During Non-Clinical Procedure	A090805	Incorrect Measurement		
16959	2025	All other reportable incidents	47263	Cardiac resynchronization therapy implantable pacemaker	CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A0908	Incorrect, Inadequate or Imprecise Result or Readings	07/05/2025	07/05/2025
							F08	Hospitalization or Prolonged Hospitalization	A0908	Incorrect, Inadequate or Imprecise Result or Readings	07/05/2025	07/05/2025
							F1905	Device Revision or Replacement	A0908	Incorrect, Inadequate or Imprecise Result or Readings	07/05/2025	07/05/2025
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A0908	Incorrect, Inadequate or Imprecise Result or Readings	07/05/2025	07/05/2025
							F08	Hospitalization or Prolonged Hospitalization	A0908	Incorrect, Inadequate or Imprecise Result or Readings	07/05/2025	07/05/2025
							F1905	Device Revision or Replacement	A0908	Incorrect, Inadequate or Imprecise Result or Readings	07/05/2025	07/05/2025
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A0908	Incorrect, Inadequate or Imprecise Result or Readings	07/05/2025	07/05/2025
							F08	Hospitalization or Prolonged Hospitalization	A0908	Incorrect, Inadequate or Imprecise Result or Readings	07/05/2025	07/05/2025
							F1905	Device Revision or Replacement	A0908	Incorrect, Inadequate or Imprecise Result or Readings	07/05/2025	07/05/2025
16960	2025	All other reportable incidents	47270	Cardiac resynchronization therapy implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E013403	Paresthesia	A070504	Premature Discharge of Battery	12/05/2025	12/05/2025
									A072202	Low Impedance	12/05/2025	12/05/2025
									A0908	Incorrect, Inadequate or Imprecise Result or Readings	12/05/2025	12/05/2025
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	12/05/2025	12/05/2025
									A072202	Low Impedance	12/05/2025	12/05/2025
									A0908	Incorrect, Inadequate or Imprecise Result or Readings	12/05/2025	12/05/2025
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	12/05/2025	12/05/2025
									A072202	Low Impedance	12/05/2025	12/05/2025
									A0908	Incorrect, Inadequate or Imprecise Result or Readings	12/05/2025	12/05/2025
16961	2025	All other reportable incidents	47270	Cardiac resynchronization therapy implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E210401	Shock from Patient Lead(s)	A070909	Oversensing		
									A071302	Inappropriate/Inadequate Shock/Stimulation		
									A090801	Signal Artifact/Noise		
									A150202	Malposition of Device		
							F08	Hospitalization or Prolonged Hospitalization	A070909	Oversensing		
									A071302	Inappropriate/Inadequate Shock/Stimulation		
									A090801	Signal Artifact/Noise		
									A150202	Malposition of Device		
							F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A070909	Oversensing		
									A071302	Inappropriate/Inadequate Shock/Stimulation		
									A090801	Signal Artifact/Noise		

									A150202	Malposition of Device		
							F1905	Device Revision or Replacement	A070909	Oversensing		
									A071302	Inappropriate/Inadequate Shock/Stimulation		
									A090801	Signal Artifact/Noise		
									A150202	Malposition of Device		
16962	2025	All other reportable incidents	17882	Physiologic-monitoring defibrillation system	CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A150102	Difficult or Delayed Activation		
							F27	Problem Identified During Non-Clinical Procedure	A150102	Difficult or Delayed Activation		
16963	2025	All other reportable incidents	17882	Physiologic-monitoring defibrillation system	CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A150102	Difficult or Delayed Activation		
							F26	No Health Consequences or Impact	A150102	Difficult or Delayed Activation		
16964	2025		47265	Dual-chamber implantable pacemaker, rate-responsive	CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A090805	Incorrect Measurement	24/04/2025	24/04/2025
							F1901	Additional Surgery	A090805	Incorrect Measurement	24/04/2025	24/04/2025
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A090805	Incorrect Measurement	24/04/2025	24/04/2025
							F1901	Additional Surgery	A090805	Incorrect Measurement	24/04/2025	24/04/2025
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A090805	Incorrect Measurement	24/04/2025	24/04/2025
							F1901	Additional Surgery	A090805	Incorrect Measurement	24/04/2025	24/04/2025
16965	2025	All other reportable incidents	47270	Cardiac resynchronization therapy implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem	18/08/2025	18/08/2025
									A070504	Premature Discharge of Battery	18/08/2025	18/08/2025
							F08	Hospitalization or Prolonged Hospitalization	A0705	Battery Problem	18/08/2025	18/08/2025
									A070504	Premature Discharge of Battery	18/08/2025	18/08/2025
							F1905	Device Revision or Replacement	A0705	Battery Problem	18/08/2025	18/08/2025
									A070504	Premature Discharge of Battery	18/08/2025	18/08/2025
16966	2025	All other reportable incidents	36106	Implantable pacemaker analyser, intraoperative	CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A11	Computer Software Problem		
							F26	No Health Consequences or Impact	A11	Computer Software Problem		
16967	2025	All other reportable incidents	47270	Cardiac resynchronization therapy implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E210401	Shock from Patient Lead(s)	A070909	Oversensing		
									A071302	Inappropriate/Inadequate Shock/Stimulation		
									A090801	Signal Artifact/Noise		
									A1301	Failure to Read Input Signal		
							F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A070909	Oversensing		
									A071302	Inappropriate/Inadequate Shock/Stimulation		
									A090801	Signal Artifact/Noise		
									A1301	Failure to Read Input Signal		
							F2203	Imaging Required	A070909	Oversensing		
									A071302	Inappropriate/Inadequate Shock/Stimulation		
									A090801	Signal Artifact/Noise		
									A1301	Failure to Read Input Signal		
16968	2025	All other reportable incidents	47263	Cardiac resynchronization therapy implantable pacemaker	CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A0203	Defective Device	05/11/2024	05/11/2024
							F24	Insufficient Health Impact Information	A0203	Defective Device	05/11/2024	05/11/2024
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A0203	Defective Device	05/11/2024	05/11/2024
							F24	Insufficient Health Impact Information	A0203	Defective Device	05/11/2024	05/11/2024
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A0203	Defective Device	05/11/2024	05/11/2024
							F24	Insufficient Health Impact Information	A0203	Defective Device	05/11/2024	05/11/2024
16969	2025	All other reportable incidents	47270	Cardiac resynchronization therapy implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E210401	Shock from Patient Lead(s)	A070504	Premature Discharge of Battery		
									A070909	Oversensing		
									A071302	Inappropriate/Inadequate Shock/Stimulation		
									A090801	Signal Artifact/Noise		
									A1301	Failure to Read Input Signal		
							F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A070504	Premature Discharge of Battery		
									A070909	Oversensing		
									A071302	Inappropriate/Inadequate Shock/Stimulation		
									A090801	Signal Artifact/Noise		
									A1301	Failure to Read Input Signal		
							F2203	Imaging Required	A070504	Premature Discharge of Battery		
									A070909	Oversensing		
									A071302	Inappropriate/Inadequate Shock/Stimulation		
									A090801	Signal Artifact/Noise		
									A1301	Failure to Read Input Signal		
16970	2025	All other reportable incidents	35852	Single-chamber implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E210401	Shock from Patient Lead(s)	A071302	Inappropriate/Inadequate Shock/Stimulation	21/05/2025	21/05/2025
							F08	Hospitalization or Prolonged Hospitalization	A071302	Inappropriate/Inadequate Shock/Stimulation	21/05/2025	21/05/2025
							F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A071302	Inappropriate/Inadequate Shock/Stimulation	21/05/2025	21/05/2025
							F1905	Device Revision or Replacement	A071302	Inappropriate/Inadequate Shock/Stimulation	21/05/2025	21/05/2025
16971	2025	All other reportable incidents	37265	Dual-chamber implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E210401	Shock from Patient Lead(s)	A0712	Pacing Problem		
									A071201	Failure to Convert Rhythm		
									A071302	Inappropriate/Inadequate Shock/Stimulation		
							F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A0712	Pacing Problem		
									A071201	Failure to Convert Rhythm		
									A071302	Inappropriate/Inadequate Shock/Stimulation		
16972	2025	All other reportable incidents	47270	Cardiac resynchronization therapy implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A0709	Device Sensing Problem		
							F12	Serious Injury/ Illness/ Impairment	A0709	Device Sensing Problem		
							F2307	Reprogramming of Device	A0709	Device Sensing Problem		
16973	2025	All other reportable incidents	35852	Single-chamber implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A072201	High Impedance		
							F26	No Health Consequences or Impact	A072201	High Impedance		
16974	2025	All other reportable incidents	17882	Physiologic-monitoring defibrillation system	CT1684	Cardiac pacemakers and associated devices	E0602	Cardiac Arrest	A071301	Failure to Deliver Shock/Stimulation		
							F02	Death	A071301	Failure to Deliver Shock/Stimulation		
16975	2025	All other reportable incidents	35852	Single-chamber implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E2104	Electric Shock	A070909	Oversensing		
									A071302	Inappropriate/Inadequate Shock/Stimulation		
							F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A070909	Oversensing		
									A071302	Inappropriate/Inadequate Shock/Stimulation		
16976	2025	All other reportable incidents	47270	Cardiac resynchronization therapy implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A180103	Device Contamination with Body Fluid	06/05/2025	06/05/2025
							F26	No Health Consequences or Impact	A180103	Device Contamination with Body Fluid	06/05/2025	06/05/2025
16977	2025	All other reportable incidents	35852	Single-chamber implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E210401	Shock from Patient Lead(s)	A071302	Inappropriate/Inadequate Shock/Stimulation		
							F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A071302	Inappropriate/Inadequate Shock/Stimulation		
							F2203	Imaging Required	A071302	Inappropriate/Inadequate Shock/Stimulation		
16978	2025	All other reportable incidents	47270	Cardiac resynchronization therapy implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E210401	Shock from Patient Lead(s)	A070909	Oversensing		
									A071302	Inappropriate/Inadequate Shock/Stimulation		
									A090801	Signal Artifact/Noise		
									A150202	Malposition of Device		
							F08	Hospitalization or Prolonged Hospitalization	A070909	Oversensing		
									A071302	Inappropriate/Inadequate Shock/Stimulation		

									A090801	Signal Artifact/Noise		
									A150202	Malposition of Device		
							F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A070909	Oversensing		
									A071302	Inappropriate/Inadequate Shock/Stimulation		
									A090801	Signal Artifact/Noise		
									A150202	Malposition of Device		
							F1905	Device Revision or Replacement	A070909	Oversensing		
									A071302	Inappropriate/Inadequate Shock/Stimulation		
									A090801	Signal Artifact/Noise		
									A150202	Malposition of Device		
16979	2025	All other reportable incidents	35852	Single-chamber implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A010402	Migration		
									A072201	High Impedance		
							F26	No Health Consequences or Impact	A010402	Migration		
									A072201	High Impedance		
16980	2025	All other reportable incidents	35852	Single-chamber implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F23	Unexpected Medical Intervention	A070504	Premature Discharge of Battery		
16981	2025	All other reportable incidents	35852	Single-chamber implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
16982	2025	All other reportable incidents	37265	Dual-chamber implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E210401	Shock from Patient Lead(s)	A0709	Device Sensing Problem		
									A071302	Inappropriate/Inadequate Shock/Stimulation		
							F08	Hospitalization or Prolonged Hospitalization	A0709	Device Sensing Problem		
									A071302	Inappropriate/Inadequate Shock/Stimulation		
							F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A0709	Device Sensing Problem		
									A071302	Inappropriate/Inadequate Shock/Stimulation		
							F1203	Life Threatening Illness or Injury	A0709	Device Sensing Problem		
									A071302	Inappropriate/Inadequate Shock/Stimulation		
							F2307	Reprogramming of Device	A0709	Device Sensing Problem		
									A071302	Inappropriate/Inadequate Shock/Stimulation		
16983	2025	All other reportable incidents	47265	Dual-chamber implantable pacemaker, rate-responsive	CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A070909	Oversensing		
									A072201	High Impedance		
							F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A070909	Oversensing		
									A072201	High Impedance		
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A070909	Oversensing		
									A072201	High Impedance		
							F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A070909	Oversensing		
									A072201	High Impedance		
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A070909	Oversensing		
									A072201	High Impedance		
							F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A070909	Oversensing		
									A072201	High Impedance		
16984	2025	All other reportable incidents	47270	Cardiac resynchronization therapy implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A0709	Device Sensing Problem		
									A070909	Oversensing		
									A070910	Undersensing		
									A090801	Signal Artifact/Noise		
									A12	Connection Problem		
									A150202	Malposition of Device		
							F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A0709	Device Sensing Problem		
									A070909	Oversensing		
									A070910	Undersensing		
									A090801	Signal Artifact/Noise		
									A12	Connection Problem		
									A150202	Malposition of Device		
							F2203	Imaging Required	A0709	Device Sensing Problem		
									A070909	Oversensing		
									A070910	Undersensing		
									A090801	Signal Artifact/Noise		
									A12	Connection Problem		
									A150202	Malposition of Device		
16985	2025	All other reportable incidents	47265	Dual-chamber implantable pacemaker, rate-responsive	CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	23/05/2025	23/05/2025
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	23/05/2025	23/05/2025
							F12	Serious Injury/ Illness/ Impairment	A070504	Premature Discharge of Battery	23/05/2025	23/05/2025
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	23/05/2025	23/05/2025
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery	23/05/2025	23/05/2025
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	23/05/2025	23/05/2025
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	23/05/2025	23/05/2025
							F12	Serious Injury/ Illness/ Impairment	A070504	Premature Discharge of Battery	23/05/2025	23/05/2025
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	23/05/2025	23/05/2025
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery	23/05/2025	23/05/2025
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	23/05/2025	23/05/2025
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	23/05/2025	23/05/2025
							F12	Serious Injury/ Illness/ Impairment	A070504	Premature Discharge of Battery	23/05/2025	23/05/2025
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	23/05/2025	23/05/2025
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery	23/05/2025	23/05/2025
16986	2025	All other reportable incidents	47270	Cardiac resynchronization therapy implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E210401	Shock from Patient Lead(s)	A071302	Inappropriate/Inadequate Shock/Stimulation		
							F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A071302	Inappropriate/Inadequate Shock/Stimulation		
							F2203	Imaging Required	A071302	Inappropriate/Inadequate Shock/Stimulation		
16987	2025	All other reportable incidents	37265	Dual-chamber implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A070909	Oversensing		
							F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A070909	Oversensing		
16988	2025		47265	Dual-chamber implantable pacemaker, rate-responsive	CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	08/05/2025	08/05/2025
							F1901	Additional Surgery	A24	No Apparent Device Problem	08/05/2025	08/05/2025
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	08/05/2025	08/05/2025
							F1901	Additional Surgery	A24	No Apparent Device Problem	08/05/2025	08/05/2025
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	08/05/2025	08/05/2025
							F1901	Additional Surgery	A24	No Apparent Device Problem	08/05/2025	08/05/2025
16989	2025	All other reportable incidents	35852	Single-chamber implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E210401	Shock from Patient Lead(s)	A0709	Device Sensing Problem		
									A071302	Inappropriate/Inadequate Shock/Stimulation		
							F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A0709	Device Sensing Problem		
									A071302	Inappropriate/Inadequate Shock/Stimulation		

							F1203	Life Threatening Illness or Injury	A0709	Device Sensing Problem		
									A071302	Inappropriate/Inadequate Shock/Stimulation		
							F2307	Reprogramming of Device	A0709	Device Sensing Problem		
									A071302	Inappropriate/Inadequate Shock/Stimulation		
16990	2025	All other reportable incidents	47265	Dual-chamber implantable pacemaker, rate-responsive	CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A1304	Telemetry Discrepancy		
							F26	No Health Consequences or Impact	A1304	Telemetry Discrepancy		
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A1304	Telemetry Discrepancy		
							F26	No Health Consequences or Impact	A1304	Telemetry Discrepancy		
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A1304	Telemetry Discrepancy		
							F26	No Health Consequences or Impact	A1304	Telemetry Discrepancy		
16991	2025	All other reportable incidents	47265	Dual-chamber implantable pacemaker, rate-responsive	CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A070101	Failure to Capture		
									A070909	Oversensing		
									A1301	Failure to Read Input Signal		
							F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A070101	Failure to Capture		
									A070909	Oversensing		
									A1301	Failure to Read Input Signal		
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A070101	Failure to Capture		
									A070909	Oversensing		
									A1301	Failure to Read Input Signal		
							F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A070101	Failure to Capture		
									A070909	Oversensing		
									A1301	Failure to Read Input Signal		
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A070101	Failure to Capture		
									A070909	Oversensing		
									A1301	Failure to Read Input Signal		
							F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A070101	Failure to Capture		
									A070909	Oversensing		
									A1301	Failure to Read Input Signal		
16992	2025		47265	Dual-chamber implantable pacemaker, rate-responsive	CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	06/05/2025	06/05/2025
							F1901	Additional Surgery	A24	No Apparent Device Problem	06/05/2025	06/05/2025
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	06/05/2025	06/05/2025
							F1901	Additional Surgery	A24	No Apparent Device Problem	06/05/2025	06/05/2025
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	06/05/2025	06/05/2025
							F1901	Additional Surgery	A24	No Apparent Device Problem	06/05/2025	06/05/2025
16993	2025	All other reportable incidents	47263	Cardiac resynchronization therapy implantable pacemaker	CT1685	Cardiac pacemakers	E0109	Convulsion/Seizure	A070909	Oversensing	15/05/2025	15/05/2025
									A0712	Pacing Problem	15/05/2025	15/05/2025
									A0908	Incorrect, Inadequate or Imprecise Result or Readings	15/05/2025	15/05/2025
							E060101	Asystole	A070909	Oversensing	15/05/2025	15/05/2025
									A0712	Pacing Problem	15/05/2025	15/05/2025
									A0908	Incorrect, Inadequate or Imprecise Result or Readings	15/05/2025	15/05/2025
							F08	Hospitalization or Prolonged Hospitalization	A070909	Oversensing	15/05/2025	15/05/2025
									A0712	Pacing Problem	15/05/2025	15/05/2025
									A0908	Incorrect, Inadequate or Imprecise Result or Readings	15/05/2025	15/05/2025
							F1001	Absence of Treatment	A070909	Oversensing	15/05/2025	15/05/2025
									A0712	Pacing Problem	15/05/2025	15/05/2025
									A0908	Incorrect, Inadequate or Imprecise Result or Readings	15/05/2025	15/05/2025
							F1905	Device Revision or Replacement	A070909	Oversensing	15/05/2025	15/05/2025
									A0712	Pacing Problem	15/05/2025	15/05/2025
									A0908	Incorrect, Inadequate or Imprecise Result or Readings	15/05/2025	15/05/2025
					CT257	Implantable cardiac pacemakers	E0109	Convulsion/Seizure	A070909	Oversensing	15/05/2025	15/05/2025
									A0712	Pacing Problem	15/05/2025	15/05/2025
									A0908	Incorrect, Inadequate or Imprecise Result or Readings	15/05/2025	15/05/2025
							E060101	Asystole	A070909	Oversensing	15/05/2025	15/05/2025
									A0712	Pacing Problem	15/05/2025	15/05/2025
									A0908	Incorrect, Inadequate or Imprecise Result or Readings	15/05/2025	15/05/2025
							F08	Hospitalization or Prolonged Hospitalization	A070909	Oversensing	15/05/2025	15/05/2025
									A0712	Pacing Problem	15/05/2025	15/05/2025
									A0908	Incorrect, Inadequate or Imprecise Result or Readings	15/05/2025	15/05/2025
							F1001	Absence of Treatment	A070909	Oversensing	15/05/2025	15/05/2025
									A0712	Pacing Problem	15/05/2025	15/05/2025
									A0908	Incorrect, Inadequate or Imprecise Result or Readings	15/05/2025	15/05/2025
							F1905	Device Revision or Replacement	A070909	Oversensing	15/05/2025	15/05/2025
									A0712	Pacing Problem	15/05/2025	15/05/2025
									A0908	Incorrect, Inadequate or Imprecise Result or Readings	15/05/2025	15/05/2025
					CT2866	Implantable cardiac electrical stimulators	E0109	Convulsion/Seizure	A070909	Oversensing	15/05/2025	15/05/2025
									A0712	Pacing Problem	15/05/2025	15/05/2025
									A0908	Incorrect, Inadequate or Imprecise Result or Readings	15/05/2025	15/05/2025
							E060101	Asystole	A070909	Oversensing	15/05/2025	15/05/2025
									A0712	Pacing Problem	15/05/2025	15/05/2025
									A0908	Incorrect, Inadequate or Imprecise Result or Readings	15/05/2025	15/05/2025
							F08	Hospitalization or Prolonged Hospitalization	A070909	Oversensing	15/05/2025	15/05/2025
									A0712	Pacing Problem	15/05/2025	15/05/2025
									A0908	Incorrect, Inadequate or Imprecise Result or Readings	15/05/2025	15/05/2025
							F1001	Absence of Treatment	A070909	Oversensing	15/05/2025	15/05/2025
									A0712	Pacing Problem	15/05/2025	15/05/2025
									A0908	Incorrect, Inadequate or Imprecise Result or Readings	15/05/2025	15/05/2025
							F1905	Device Revision or Replacement	A070909	Oversensing	15/05/2025	15/05/2025

									A0712	Pacing Problem	15/05/2025	15/05/2025
									A0908	Incorrect, Inadequate or Imprecise Result or Readings	15/05/2025	15/05/2025
16994	2025		47265	Dual-chamber implantable pacemaker, rate-responsive	CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	04/04/2025	04/04/2025
							F1901	Additional Surgery	A24	No Apparent Device Problem	04/04/2025	04/04/2025
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	04/04/2025	04/04/2025
							F1901	Additional Surgery	A24	No Apparent Device Problem	04/04/2025	04/04/2025
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	04/04/2025	04/04/2025
							F1901	Additional Surgery	A24	No Apparent Device Problem	04/04/2025	04/04/2025
16995	2025	All other reportable incidents	63189	Implantable pacemaker battery	CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070102	High Capture Threshold		
							F26	No Health Consequences or Impact	A070102	High Capture Threshold		
16996	2025	All other reportable incidents	47270	Cardiac resynchronization therapy implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E011903	Syncope/Fainting	A0709	Device Sensing Problem		
									A071301	Failure to Deliver Shock/Stimulation		
							E060109	Tachycardia	A0709	Device Sensing Problem		
									A071301	Failure to Deliver Shock/Stimulation		
							F1001	Absence of Treatment	A0709	Device Sensing Problem		
									A071301	Failure to Deliver Shock/Stimulation		
							F1203	Life Threatening Illness or Injury	A0709	Device Sensing Problem		
									A071301	Failure to Deliver Shock/Stimulation		
							F2307	Reprogramming of Device	A0709	Device Sensing Problem		
									A071301	Failure to Deliver Shock/Stimulation		
16997	2025	All other reportable incidents	35852	Single-chamber implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A07	Electrical /Electronic Property Problem	13/05/2025	13/05/2025
							F08	Hospitalization or Prolonged Hospitalization	A07	Electrical /Electronic Property Problem	13/05/2025	13/05/2025
							F12	Serious Injury/ Illness/ Impairment	A07	Electrical /Electronic Property Problem	13/05/2025	13/05/2025
							F1905	Device Revision or Replacement	A07	Electrical /Electronic Property Problem	13/05/2025	13/05/2025
16998	2025	All other reportable incidents	35852	Single-chamber implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery		
							F12	Serious Injury/ Illness/ Impairment	A070504	Premature Discharge of Battery		
							F1903	Device Explantation	A070504	Premature Discharge of Battery		
16999	2025	All other reportable incidents	37265	Dual-chamber implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A070602	Delayed Charge Time		
									A070910	Undersensing		
									A072201	High Impedance		
									A1301	Failure to Read Input Signal		
							F15	Recognized Device or Procedural Complication	A070602	Delayed Charge Time		
									A070910	Undersensing		
									A072201	High Impedance		
									A1301	Failure to Read Input Signal		
17000	2025	All other reportable incidents	17882	Physiologic-monitoring defibrillation system	CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A0401	Break		
							F27	Problem Identified During Non-Clinical Procedure	A0401	Break		
17001	2025	All other reportable incidents	17882	Physiologic-monitoring defibrillation system	CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A0401	Break		
							F27	Problem Identified During Non-Clinical Procedure	A0401	Break		
17002	2025	All other reportable incidents	17882	Physiologic-monitoring defibrillation system	CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070903	Failure to Analyze Signal		
							F26	No Health Consequences or Impact	A070903	Failure to Analyze Signal		
17003	2025	All other reportable incidents	17882	Physiologic-monitoring defibrillation system	CT1684	Cardiac pacemakers and associated devices	E1704	Burn(s)	A0101	Patient-Device Incompatibility		
							F02	Death	A0101	Patient-Device Incompatibility		
17004	2025	All other reportable incidents	37265	Dual-chamber implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A0709	Device Sensing Problem		
							F12	Serious Injury/ Illness/ Impairment	A0709	Device Sensing Problem		
							F2307	Reprogramming of Device	A0709	Device Sensing Problem		
17005	2025	All other reportable incidents	47267	Single-chamber implantable pacemaker, rate-responsive	CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
17006	2025	All other reportable incidents	47265	Dual-chamber implantable pacemaker, rate-responsive	CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	27/05/2025	27/05/2025
							F1901	Additional Surgery	A070504	Premature Discharge of Battery	27/05/2025	27/05/2025
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	27/05/2025	27/05/2025
							F1901	Additional Surgery	A070504	Premature Discharge of Battery	27/05/2025	27/05/2025
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	27/05/2025	27/05/2025
							F1901	Additional Surgery	A070504	Premature Discharge of Battery	27/05/2025	27/05/2025
17007	2025	All other reportable incidents	47205	Cardiac pulse generator programmer	CT1684	Cardiac pacemakers and associated devices	E060102	Atrial Fibrillation	A09	Output Problem		
									A0902	Display or Visual Feedback Problem		
									A13	Communication or Transmission Problem		
							E060109	Tachycardia	A09	Output Problem		
									A0902	Display or Visual Feedback Problem		
									A13	Communication or Transmission Problem		
							F1203	Life Threatening Illness or Injury	A09	Output Problem		
									A0902	Display or Visual Feedback Problem		
									A13	Communication or Transmission Problem		
							F23	Unexpected Medical Intervention	A09	Output Problem		
									A0902	Display or Visual Feedback Problem		
									A13	Communication or Transmission Problem		
17008	2025	All other reportable incidents	47265	Dual-chamber implantable pacemaker, rate-responsive	CT1685	Cardiac pacemakers	E0112	Dizziness	A05	Mechanical Problem		
									A0701	Capturing Problem		
									A070101	Failure to Capture		
									A0722	Impedance Problem		
									A1205	Loose or Intermittent Connection		
							F08	Hospitalization or Prolonged Hospitalization	A05	Mechanical Problem		
									A0701	Capturing Problem		
									A070101	Failure to Capture		
									A0722	Impedance Problem		
									A1205	Loose or Intermittent Connection		
							F12	Serious Injury/ Illness/ Impairment	A05	Mechanical Problem		
									A0701	Capturing Problem		
									A070101	Failure to Capture		
									A0722	Impedance Problem		
									A1205	Loose or Intermittent Connection		
							F1905	Device Revision or Replacement	A05	Mechanical Problem		

									A0701	Capturing Problem		
									A070101	Failure to Capture		
									A0722	Impedance Problem		
									A1205	Loose or Intermittent Connection		
					CT257	Implantable cardiac pacemakers	E0112	Dizziness	A05	Mechanical Problem		
									A0701	Capturing Problem		
									A070101	Failure to Capture		
									A0722	Impedance Problem		
									A1205	Loose or Intermittent Connection		
							F08	Hospitalization or Prolonged Hospitalization	A05	Mechanical Problem		
									A0701	Capturing Problem		
									A070101	Failure to Capture		
									A0722	Impedance Problem		
									A1205	Loose or Intermittent Connection		
							F12	Serious Injury/ Illness/ Impairment	A05	Mechanical Problem		
									A0701	Capturing Problem		
									A070101	Failure to Capture		
									A0722	Impedance Problem		
									A1205	Loose or Intermittent Connection		
							F1905	Device Revision or Replacement	A05	Mechanical Problem		
									A0701	Capturing Problem		
									A070101	Failure to Capture		
									A0722	Impedance Problem		
									A1205	Loose or Intermittent Connection		
					CT2866	Implantable cardiac electrical stimulators	E0112	Dizziness	A05	Mechanical Problem		
									A0701	Capturing Problem		
									A070101	Failure to Capture		
									A0722	Impedance Problem		
									A1205	Loose or Intermittent Connection		
							F08	Hospitalization or Prolonged Hospitalization	A05	Mechanical Problem		
									A0701	Capturing Problem		
									A070101	Failure to Capture		
									A0722	Impedance Problem		
									A1205	Loose or Intermittent Connection		
							F12	Serious Injury/ Illness/ Impairment	A05	Mechanical Problem		
									A0701	Capturing Problem		
									A070101	Failure to Capture		
									A0722	Impedance Problem		
									A1205	Loose or Intermittent Connection		
							F1905	Device Revision or Replacement	A05	Mechanical Problem		
									A0701	Capturing Problem		
									A070101	Failure to Capture		
									A0722	Impedance Problem		
									A1205	Loose or Intermittent Connection		
17009	2025		47265	Dual-chamber implantable pacemaker, rate-responsive	CT1685	Cardiac pacemakers	E011902	Presyncope	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
							E060104	Bradycardia	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
							E060106	Heart Block	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
							F08	Hospitalization or Prolonged Hospitalization	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
							F12	Serious Injury/ Illness/ Impairment	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
					CT257	Implantable cardiac pacemakers	E011902	Presyncope	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
							E060104	Bradycardia	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
							E060106	Heart Block	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
							F08	Hospitalization or Prolonged Hospitalization	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
							F12	Serious Injury/ Illness/ Impairment	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
					CT2866	Implantable cardiac electrical stimulators	E011902	Presyncope	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
							E060104	Bradycardia	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
							E060106	Heart Block	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
							F08	Hospitalization or Prolonged Hospitalization	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
							F12	Serious Injury/ Illness/ Impairment	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
17010	2025	All other reportable incidents	47267	Single-chamber implantable pacemaker, rate-responsive	CT1685	Cardiac pacemakers	E0601	Arrhythmia	A24	No Apparent Device Problem	01/05/2025	01/05/2025
							E0606	Cardiomyopathy	A24	No Apparent Device Problem	01/05/2025	01/05/2025
							E0611	Heart Failure/Congestive Heart Failure	A24	No Apparent Device Problem	01/05/2025	01/05/2025
							E2338	Swelling/ Edema	A24	No Apparent Device Problem	01/05/2025	01/05/2025
							F08	Hospitalization or Prolonged Hospitalization	A24	No Apparent Device Problem	01/05/2025	01/05/2025
							F12	Serious Injury/ Illness/ Impairment	A24	No Apparent Device Problem	01/05/2025	01/05/2025
							F1905	Device Revision or Replacement	A24	No Apparent Device Problem	01/05/2025	01/05/2025
					CT257	Implantable cardiac pacemakers	E0601	Arrhythmia	A24	No Apparent Device Problem	01/05/2025	01/05/2025
							E0606	Cardiomyopathy	A24	No Apparent Device Problem	01/05/2025	01/05/2025
							E0611	Heart Failure/Congestive Heart Failure	A24	No Apparent Device Problem	01/05/2025	01/05/2025
							E2338	Swelling/ Edema	A24	No Apparent Device Problem	01/05/2025	01/05/2025
							F08	Hospitalization or Prolonged Hospitalization	A24	No Apparent Device Problem	01/05/2025	01/05/2025
							F12	Serious Injury/ Illness/ Impairment	A24	No Apparent Device Problem	01/05/2025	01/05/2025
							F1905	Device Revision or Replacement	A24	No Apparent Device Problem	01/05/2025	01/05/2025

					CT2866	Implantable cardiac electrical stimulators	E0601	Arrhythmia	A24	No Apparent Device Problem	01/05/2025	01/05/2025
							E0606	Cardiomyopathy	A24	No Apparent Device Problem	01/05/2025	01/05/2025
							E0611	Heart Failure/Congestive Heart Failure	A24	No Apparent Device Problem	01/05/2025	01/05/2025
							E2338	Swelling/ Edema	A24	No Apparent Device Problem	01/05/2025	01/05/2025
							F08	Hospitalization or Prolonged Hospitalization	A24	No Apparent Device Problem	01/05/2025	01/05/2025
							F12	Serious Injury/ Illness/ Impairment	A24	No Apparent Device Problem	01/05/2025	01/05/2025
							F1905	Device Revision or Replacement	A24	No Apparent Device Problem	01/05/2025	01/05/2025
17011	2025	All other reportable incidents	37265	Dual-chamber implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A070909	Oversensing		
									A090801	Signal Artifact/Noise		
							F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A070909	Oversensing		
									A090801	Signal Artifact/Noise		
17012	2025	All other reportable incidents	47265	Dual-chamber implantable pacemaker, rate-responsive	CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A0908	Incorrect, Inadequate or Imprecise Result or Readings		
							F26	No Health Consequences or Impact	A0908	Incorrect, Inadequate or Imprecise Result or Readings		
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A0908	Incorrect, Inadequate or Imprecise Result or Readings		
							F26	No Health Consequences or Impact	A0908	Incorrect, Inadequate or Imprecise Result or Readings		
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A0908	Incorrect, Inadequate or Imprecise Result or Readings		
							F26	No Health Consequences or Impact	A0908	Incorrect, Inadequate or Imprecise Result or Readings		
17013	2025	All other reportable incidents	47270	Cardiac resynchronization therapy implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E210401	Shock from Patient Lead(s)	A070909	Oversensing		
									A071302	Inappropriate/Inadequate Shock/Stimulation		
									A1301	Failure to Read Input Signal		
							F08	Hospitalization or Prolonged Hospitalization	A070909	Oversensing		
									A071302	Inappropriate/Inadequate Shock/Stimulation		
									A1301	Failure to Read Input Signal		
							F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A070909	Oversensing		
									A071302	Inappropriate/Inadequate Shock/Stimulation		
									A1301	Failure to Read Input Signal		
17014	2025	All other reportable incidents	47270	Cardiac resynchronization therapy implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A070909	Oversensing		
									A072001	Electromagnetic Interference		
							F12	Serious Injury/ Illness/ Impairment	A070909	Oversensing		
									A072001	Electromagnetic Interference		
							F2307	Reprogramming of Device	A070909	Oversensing		
									A072001	Electromagnetic Interference		
17015	2025	All other reportable incidents	47263	Cardiac resynchronization therapy implantable pacemaker	CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A0908	Incorrect, Inadequate or Imprecise Result or Readings		
							F26	No Health Consequences or Impact	A0908	Incorrect, Inadequate or Imprecise Result or Readings		
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A0908	Incorrect, Inadequate or Imprecise Result or Readings		
							F26	No Health Consequences or Impact	A0908	Incorrect, Inadequate or Imprecise Result or Readings		
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A0908	Incorrect, Inadequate or Imprecise Result or Readings		
							F26	No Health Consequences or Impact	A0908	Incorrect, Inadequate or Imprecise Result or Readings		
17016	2025	All other reportable incidents	47205	Cardiac pulse generator programmer	CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A13	Communication or Transmission Problem		
							F26	No Health Consequences or Impact	A13	Communication or Transmission Problem		
17017	2025	All other reportable incidents	35852	Single-chamber implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E210401	Shock from Patient Lead(s)	A0709	Device Sensing Problem		
									A071302	Inappropriate/Inadequate Shock/Stimulation		
							F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A0709	Device Sensing Problem		
									A071302	Inappropriate/Inadequate Shock/Stimulation		
							F1203	Life Threatening Illness or Injury	A0709	Device Sensing Problem		
									A071302	Inappropriate/Inadequate Shock/Stimulation		
							F2307	Reprogramming of Device	A0709	Device Sensing Problem		
									A071302	Inappropriate/Inadequate Shock/Stimulation		
17018	2025	All other reportable incidents	47270	Cardiac resynchronization therapy implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E060110	Ventricular Fibrillation	A0709	Device Sensing Problem		
									A0712	Pacing Problem		
									A0713	Defibrillation/Stimulation Problem		
									A071301	Failure to Deliver Shock/Stimulation		
							F05	Delay to Treatment/ Therapy	A0709	Device Sensing Problem		
									A0712	Pacing Problem		
									A0713	Defibrillation/Stimulation Problem		
									A071301	Failure to Deliver Shock/Stimulation		
							F1001	Absence of Treatment	A0709	Device Sensing Problem		
									A0712	Pacing Problem		
									A0713	Defibrillation/Stimulation Problem		
									A071301	Failure to Deliver Shock/Stimulation		
							F1203	Life Threatening Illness or Injury	A0709	Device Sensing Problem		
									A0712	Pacing Problem		
									A0713	Defibrillation/Stimulation Problem		
									A071301	Failure to Deliver Shock/Stimulation		
17019	2025	All other reportable incidents	47270	Cardiac resynchronization therapy implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E060109	Tachycardia	A071301	Failure to Deliver Shock/Stimulation		
									A13	Communication or Transmission Problem		
							F1001	Absence of Treatment	A071301	Failure to Deliver Shock/Stimulation		
									A13	Communication or Transmission Problem		
							F1203	Life Threatening Illness or Injury	A071301	Failure to Deliver Shock/Stimulation		
									A13	Communication or Transmission Problem		
17020	2025	All other reportable incidents	47270	Cardiac resynchronization therapy implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A070909	Oversensing		
							F26	No Health Consequences or Impact	A070909	Oversensing		
17021	2025	All other reportable incidents	47267	Single-chamber implantable pacemaker, rate-responsive	CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A09	Output Problem		
							F26	No Health Consequences or Impact	A09	Output Problem		
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A09	Output Problem		
							F26	No Health Consequences or Impact	A09	Output Problem		
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A09	Output Problem		
							F26	No Health Consequences or Impact	A09	Output Problem		
17022	2025	All other reportable incidents	47265	Dual-chamber implantable pacemaker, rate-	CT1685	Cardiac pacemakers	E060109	Tachycardia	A0712	Pacing Problem		

				responsive			E2402	Appropriate Clinical Signs, Symptoms and Conditions Term/Code not Available	A0712	Pacing Problem		
							F11	Non-Serious Injury/ Illness/ Impairment	A0712	Pacing Problem		
							F28	Appropriate Health Impact Term/Code not Available	A0712	Pacing Problem		
					CT257	Implantable cardiac pacemakers	E060109	Tachycardia	A0712	Pacing Problem		
							E2402	Appropriate Clinical Signs, Symptoms and Conditions Term/Code not Available	A0712	Pacing Problem		
							F11	Non-Serious Injury/ Illness/ Impairment	A0712	Pacing Problem		
							F28	Appropriate Health Impact Term/Code not Available	A0712	Pacing Problem		
					CT2866	Implantable cardiac electrical stimulators	E060109	Tachycardia	A0712	Pacing Problem		
							E2402	Appropriate Clinical Signs, Symptoms and Conditions Term/Code not Available	A0712	Pacing Problem		
							F11	Non-Serious Injury/ Illness/ Impairment	A0712	Pacing Problem		
							F28	Appropriate Health Impact Term/Code not Available	A0712	Pacing Problem		
17023	2025	All other reportable incidents	37265	Dual-chamber implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A070909	Oversensing		
							F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A070909	Oversensing		
17024	2025	All other reportable incidents	47267	Single-chamber implantable pacemaker, rate-responsive	CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A05	Mechanical Problem	07/04/2025	07/04/2025
							F08	Hospitalization or Prolonged Hospitalization	A05	Mechanical Problem	07/04/2025	07/04/2025
							F1905	Device Revision or Replacement	A05	Mechanical Problem	07/04/2025	07/04/2025
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A05	Mechanical Problem	07/04/2025	07/04/2025
							F08	Hospitalization or Prolonged Hospitalization	A05	Mechanical Problem	07/04/2025	07/04/2025
							F1905	Device Revision or Replacement	A05	Mechanical Problem	07/04/2025	07/04/2025
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A05	Mechanical Problem	07/04/2025	07/04/2025
							F08	Hospitalization or Prolonged Hospitalization	A05	Mechanical Problem	07/04/2025	07/04/2025
							F1905	Device Revision or Replacement	A05	Mechanical Problem	07/04/2025	07/04/2025
17025	2025	All other reportable incidents	47267	Single-chamber implantable pacemaker, rate-responsive	CT1685	Cardiac pacemakers	E060110	Ventricular Fibrillation	A0712	Pacing Problem	22/05/2025	22/05/2025
							F08	Hospitalization or Prolonged Hospitalization	A0712	Pacing Problem	22/05/2025	22/05/2025
							F1203	Life Threatening Illness or Injury	A0712	Pacing Problem	22/05/2025	22/05/2025
							F1905	Device Revision or Replacement	A0712	Pacing Problem	22/05/2025	22/05/2025
							F2307	Reprogramming of Device	A0712	Pacing Problem	22/05/2025	22/05/2025
					CT257	Implantable cardiac pacemakers	E060110	Ventricular Fibrillation	A0712	Pacing Problem	22/05/2025	22/05/2025
							F08	Hospitalization or Prolonged Hospitalization	A0712	Pacing Problem	22/05/2025	22/05/2025
							F1203	Life Threatening Illness or Injury	A0712	Pacing Problem	22/05/2025	22/05/2025
							F1905	Device Revision or Replacement	A0712	Pacing Problem	22/05/2025	22/05/2025
							F2307	Reprogramming of Device	A0712	Pacing Problem	22/05/2025	22/05/2025
					CT2866	Implantable cardiac electrical stimulators	E060110	Ventricular Fibrillation	A0712	Pacing Problem	22/05/2025	22/05/2025
							F08	Hospitalization or Prolonged Hospitalization	A0712	Pacing Problem	22/05/2025	22/05/2025
							F1203	Life Threatening Illness or Injury	A0712	Pacing Problem	22/05/2025	22/05/2025
							F1905	Device Revision or Replacement	A0712	Pacing Problem	22/05/2025	22/05/2025
							F2307	Reprogramming of Device	A0712	Pacing Problem	22/05/2025	22/05/2025
17026	2025		17882	Physiologic-monitoring defibrillation system	CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A0203	Defective Device		
									A072102	Electrical Shorting		
							F26	No Health Consequences or Impact	A0203	Defective Device		
									A072102	Electrical Shorting		
17027	2025	All other reportable incidents	47270	Cardiac resynchronization therapy implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E210401	Shock from Patient Lead(s)	A070909	Oversensing		
									A071302	Inappropriate/Inadequate Shock/Stimulation		
									A1301	Failure to Read Input Signal		
							F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A070909	Oversensing		
									A071302	Inappropriate/Inadequate Shock/Stimulation		
									A1301	Failure to Read Input Signal		
17028	2025	All other reportable incidents	47263	Cardiac resynchronization therapy implantable pacemaker	CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A160105	False Alarm		
							F26	No Health Consequences or Impact	A160105	False Alarm		
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A160105	False Alarm		
							F26	No Health Consequences or Impact	A160105	False Alarm		
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A160105	False Alarm		
							F26	No Health Consequences or Impact	A160105	False Alarm		
17029	2025	All other reportable incidents	47263	Cardiac resynchronization therapy implantable pacemaker	CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
17030	2025	All other reportable incidents	17882	Physiologic-monitoring defibrillation system	CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070803	Failure to Power Up		
							F27	Problem Identified During Non-Clinical Procedure	A070803	Failure to Power Up		
17031	2025	Death	60789	Intracardiac pacemaker	CT1685	Cardiac pacemakers	E0506	Hemorrhage/Blood Loss/Bleeding	A24	No Apparent Device Problem		
							E051101	Great Vessel Perforation	A24	No Apparent Device Problem		
							E0602	Cardiac Arrest	A24	No Apparent Device Problem		
							F02	Death	A24	No Apparent Device Problem		
					CT257	Implantable cardiac pacemakers	E0506	Hemorrhage/Blood Loss/Bleeding	A24	No Apparent Device Problem		
							E051101	Great Vessel Perforation	A24	No Apparent Device Problem		
							E0602	Cardiac Arrest	A24	No Apparent Device Problem		
							F02	Death	A24	No Apparent Device Problem		
					CT2818	Catheterization/Cannulation/Fluid path devices	E0506	Hemorrhage/Blood Loss/Bleeding	A24	No Apparent Device Problem		
							E051101	Great Vessel Perforation	A24	No Apparent Device Problem		
							E0602	Cardiac Arrest	A24	No Apparent Device Problem		
							F02	Death	A24	No Apparent Device Problem		
					CT2866	Implantable cardiac electrical stimulators	E0506	Hemorrhage/Blood Loss/Bleeding	A24	No Apparent Device Problem		
							E051101	Great Vessel Perforation	A24	No Apparent Device Problem		
							E0602	Cardiac Arrest	A24	No Apparent Device Problem		
							F02	Death	A24	No Apparent Device Problem		
17032	2025	All other reportable incidents	47263	Cardiac resynchronization therapy implantable pacemaker	CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A05	Mechanical Problem		
							F12	Serious Injury/ Illness/ Impairment	A05	Mechanical Problem		
							F2307	Reprogramming of Device	A05	Mechanical Problem		
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A05	Mechanical Problem		
							F12	Serious Injury/ Illness/ Impairment	A05	Mechanical Problem		
							F2307	Reprogramming of Device	A05	Mechanical Problem		
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A05	Mechanical Problem		
							F12	Serious Injury/ Illness/ Impairment	A05	Mechanical Problem		

							F2307	Reprogramming of Device	A05	Mechanical Problem		
17033	2025	All other reportable incidents	35852	Single-chamber implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A070909	Oversensing		
									A072001	Electromagnetic Interference		
							F12	Serious Injury/ Illness/ Impairment	A070909	Oversensing		
									A072001	Electromagnetic Interference		
							F2307	Reprogramming of Device	A070909	Oversensing		
									A072001	Electromagnetic Interference		
17034	2025	All other reportable incidents	47265	Dual-chamber implantable pacemaker, rate-responsive			E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	17/09/2025	17/09/2025
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	17/09/2025	17/09/2025
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	17/09/2025	17/09/2025
17035	2025	All other reportable incidents	35852	Single-chamber implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A070909	Oversensing		
							F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A070909	Oversensing		
17036	2025	All other reportable incidents			CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A04	Material Integrity Problem	24/01/2025	24/01/2025
							F1908	Prolonged Surgery	A04	Material Integrity Problem	24/01/2025	24/01/2025
17037	2025	All other reportable incidents	47270	Cardiac resynchronization therapy implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A070910	Undersensing		
							F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A070910	Undersensing		
17038	2025	All other reportable incidents	35852	Single-chamber implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A070102	High Capture Threshold		
									A070504	Premature Discharge of Battery		
									A070910	Undersensing		
									A1301	Failure to Read Input Signal		
							F15	Recognized Device or Procedural Complication	A070102	High Capture Threshold		
									A070504	Premature Discharge of Battery		
									A070910	Undersensing		
									A1301	Failure to Read Input Signal		
17039	2025	All other reportable incidents	47263	Cardiac resynchronization therapy implantable pacemaker	CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A070101	Failure to Capture		
									A070102	High Capture Threshold		
									A1301	Failure to Read Input Signal		
							F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A070101	Failure to Capture		
									A070102	High Capture Threshold		
									A1301	Failure to Read Input Signal		
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A070101	Failure to Capture		
									A070102	High Capture Threshold		
									A1301	Failure to Read Input Signal		
							F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A070101	Failure to Capture		
									A070102	High Capture Threshold		
									A1301	Failure to Read Input Signal		
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A070101	Failure to Capture		
									A070102	High Capture Threshold		
									A1301	Failure to Read Input Signal		
							F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A070101	Failure to Capture		
									A070102	High Capture Threshold		
									A1301	Failure to Read Input Signal		
17040	2025	All other reportable incidents	47263	Cardiac resynchronization therapy implantable pacemaker	CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A1205	Loose or Intermittent Connection		
							F08	Hospitalization or Prolonged Hospitalization	A1205	Loose or Intermittent Connection		
							F12	Serious Injury/ Illness/ Impairment	A1205	Loose or Intermittent Connection		
							F1905	Device Revision or Replacement	A1205	Loose or Intermittent Connection		
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A1205	Loose or Intermittent Connection		
							F08	Hospitalization or Prolonged Hospitalization	A1205	Loose or Intermittent Connection		
							F12	Serious Injury/ Illness/ Impairment	A1205	Loose or Intermittent Connection		
							F1905	Device Revision or Replacement	A1205	Loose or Intermittent Connection		
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A1205	Loose or Intermittent Connection		
							F08	Hospitalization or Prolonged Hospitalization	A1205	Loose or Intermittent Connection		
							F12	Serious Injury/ Illness/ Impairment	A1205	Loose or Intermittent Connection		
							F1905	Device Revision or Replacement	A1205	Loose or Intermittent Connection		
17041	2025	All other reportable incidents	47270	Cardiac resynchronization therapy implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A0908	Incorrect, Inadequate or Imprecise Result or Readings		
									A1107	Data Problem		
									A160503	Inappropriate or Unexpected Reset		
							F26	No Health Consequences or Impact	A0908	Incorrect, Inadequate or Imprecise Result or Readings		
									A1107	Data Problem		
									A160503	Inappropriate or Unexpected Reset		
17042	2025	All other reportable incidents	47270	Cardiac resynchronization therapy implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A070101	Failure to Capture		
									A070909	Oversensing		
									A090801	Signal Artifact/Noise		
							F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A070101	Failure to Capture		
									A070909	Oversensing		
									A090801	Signal Artifact/Noise		
							F23	Unexpected Medical Intervention	A070101	Failure to Capture		
									A070909	Oversensing		
									A090801	Signal Artifact/Noise		
17043	2025	All other reportable incidents	47265	Dual-chamber implantable pacemaker, rate-responsive	CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A0908	Incorrect, Inadequate or Imprecise Result or Readings		
							F26	No Health Consequences or Impact	A0908	Incorrect, Inadequate or Imprecise Result or Readings		
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A0908	Incorrect, Inadequate or Imprecise Result or Readings		
							F26	No Health Consequences or Impact	A0908	Incorrect, Inadequate or Imprecise Result or Readings		
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A0908	Incorrect, Inadequate or Imprecise Result or Readings		
							F26	No Health Consequences or Impact	A0908	Incorrect, Inadequate or Imprecise Result or Readings		
17044	2025	All other reportable incidents	47205	Cardiac pulse generator programmer	CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A11	Computer Software Problem		
									A13	Communication or Transmission Problem		
							F26	No Health Consequences or Impact	A11	Computer Software Problem		
									A13	Communication or Transmission Problem		
17045	2025	All other reportable incidents	47265	Dual-chamber implantable pacemaker, rate-responsive	CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem		
									A110502	Unauthorized Access to Computer System		
									A23	Use of Device Problem		

							F26	No Health Consequences or Impact	A0705	Battery Problem		
									A110502	Unauthorized Access to Computer System		
									A23	Use of Device Problem		
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem		
									A110502	Unauthorized Access to Computer System		
									A23	Use of Device Problem		
							F26	No Health Consequences or Impact	A0705	Battery Problem		
									A110502	Unauthorized Access to Computer System		
									A23	Use of Device Problem		
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem		
									A110502	Unauthorized Access to Computer System		
									A23	Use of Device Problem		
							F26	No Health Consequences or Impact	A0705	Battery Problem		
									A110502	Unauthorized Access to Computer System		
									A23	Use of Device Problem		
17046	2025	All other reportable incidents	47205	Cardiac pulse generator programmer	CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A13	Communication or Transmission Problem		
							F26	No Health Consequences or Impact	A13	Communication or Transmission Problem		
17047	2025	All other reportable incidents	47263	Cardiac resynchronization therapy implantable pacemaker	CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem		
									A0908	Incorrect, Inadequate or Imprecise Result or Readings		
							F08	Hospitalization or Prolonged Hospitalization	A0705	Battery Problem		
									A0908	Incorrect, Inadequate or Imprecise Result or Readings		
							F1905	Device Revision or Replacement	A0705	Battery Problem		
									A0908	Incorrect, Inadequate or Imprecise Result or Readings		
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem		
									A0908	Incorrect, Inadequate or Imprecise Result or Readings		
							F08	Hospitalization or Prolonged Hospitalization	A0705	Battery Problem		
									A0908	Incorrect, Inadequate or Imprecise Result or Readings		
							F1905	Device Revision or Replacement	A0705	Battery Problem		
									A0908	Incorrect, Inadequate or Imprecise Result or Readings		
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem		
									A0908	Incorrect, Inadequate or Imprecise Result or Readings		
							F08	Hospitalization or Prolonged Hospitalization	A0705	Battery Problem		
									A0908	Incorrect, Inadequate or Imprecise Result or Readings		
							F1905	Device Revision or Replacement	A0705	Battery Problem		
									A0908	Incorrect, Inadequate or Imprecise Result or Readings		
17048	2025	All other reportable incidents	47265	Dual-chamber implantable pacemaker, rate-responsive	CT1685	Cardiac pacemakers	E060104	Bradycardia	A070504	Premature Discharge of Battery	23/05/2025	23/05/2025
									A0712	Pacing Problem	23/05/2025	23/05/2025
									A1304	Telemetry Discrepancy	23/05/2025	23/05/2025
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	23/05/2025	23/05/2025
									A0712	Pacing Problem	23/05/2025	23/05/2025
									A1304	Telemetry Discrepancy	23/05/2025	23/05/2025
							F1001	Absence of Treatment	A070504	Premature Discharge of Battery	23/05/2025	23/05/2025
									A0712	Pacing Problem	23/05/2025	23/05/2025
									A1304	Telemetry Discrepancy	23/05/2025	23/05/2025
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	23/05/2025	23/05/2025
									A0712	Pacing Problem	23/05/2025	23/05/2025
									A1304	Telemetry Discrepancy	23/05/2025	23/05/2025
					CT257	Implantable cardiac pacemakers	E060104	Bradycardia	A070504	Premature Discharge of Battery	23/05/2025	23/05/2025
									A0712	Pacing Problem	23/05/2025	23/05/2025
									A1304	Telemetry Discrepancy	23/05/2025	23/05/2025
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	23/05/2025	23/05/2025
									A0712	Pacing Problem	23/05/2025	23/05/2025
									A1304	Telemetry Discrepancy	23/05/2025	23/05/2025
							F1001	Absence of Treatment	A070504	Premature Discharge of Battery	23/05/2025	23/05/2025
									A0712	Pacing Problem	23/05/2025	23/05/2025
									A1304	Telemetry Discrepancy	23/05/2025	23/05/2025
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	23/05/2025	23/05/2025
									A0712	Pacing Problem	23/05/2025	23/05/2025
									A1304	Telemetry Discrepancy	23/05/2025	23/05/2025
					CT2866	Implantable cardiac electrical stimulators	E060104	Bradycardia	A070504	Premature Discharge of Battery	23/05/2025	23/05/2025
									A0712	Pacing Problem	23/05/2025	23/05/2025
									A1304	Telemetry Discrepancy	23/05/2025	23/05/2025
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	23/05/2025	23/05/2025
									A0712	Pacing Problem	23/05/2025	23/05/2025
									A1304	Telemetry Discrepancy	23/05/2025	23/05/2025
							F1001	Absence of Treatment	A070504	Premature Discharge of Battery	23/05/2025	23/05/2025
									A0712	Pacing Problem	23/05/2025	23/05/2025
									A1304	Telemetry Discrepancy	23/05/2025	23/05/2025
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	23/05/2025	23/05/2025
									A0712	Pacing Problem	23/05/2025	23/05/2025
									A1304	Telemetry Discrepancy	23/05/2025	23/05/2025
17049	2025	All other reportable incidents	35852	Single-chamber implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A0709	Device Sensing Problem		
							F12	Serious Injury/ Illness/ Impairment	A0709	Device Sensing Problem		
							F2307	Reprogramming of Device	A0709	Device Sensing Problem		
17050	2025	All other reportable incidents	47270	Cardiac resynchronization therapy implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E011903	Syncope/Fainting	A072001	Electromagnetic Interference		
									A090801	Signal Artifact/Noise		
							F1203	Life Threatening Illness or Injury	A072001	Electromagnetic Interference		
									A090801	Signal Artifact/Noise		
17051	2025	All other reportable incidents	17882	Physiologic-monitoring defibrillation system	CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A0405	Degraded		
							F27	Problem Identified During Non-Clinical Procedure	A0405	Degraded		
17052	2025	All other reportable incidents	63188	Left-ventricular wireless pacing system	CT145	Cardiovascular devices			A0708	Power Problem		

				sensor/transmitter	CT1577	Wireless monitoring system transmitters			A0708	Power Problem		
					CT1684	Cardiac pacemakers and associated devices			A0708	Power Problem		
					CT2727	Active implant devices/ancillaries			A0708	Power Problem		
					CT2750	Implantable cardiac electrical stimulators and associated devices			A0708	Power Problem		
					CT2885	Physiological monitoring/measuring system transmitters/receivers/transceivers			A0708	Power Problem		
17053	2025		35224	External pacemaker, epicardial pacing	CT1685	Cardiac pacemakers	E060104	Bradycardia	A160602	Premature End-of-Life Indicator		
							E060106	Heart Block	A160602	Premature End-of-Life Indicator		
							E0717	Dyspnea	A160602	Premature End-of-Life Indicator		
							F01	Change in Therapeutic Response	A160602	Premature End-of-Life Indicator		
							F1905	Device Revision or Replacement	A160602	Premature End-of-Life Indicator		
17054	2025	All other reportable incidents	17882	Physiologic-monitoring defibrillation system	CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A150102	Difficult or Delayed Activation		
							F27	Problem Identified During Non-Clinical Procedure	A150102	Difficult or Delayed Activation		
17055	2025	All other reportable incidents	47270	Cardiac resynchronization therapy implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
									A0908	Incorrect, Inadequate or Imprecise Result or Readings		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
									A0908	Incorrect, Inadequate or Imprecise Result or Readings		
17056	2025	All other reportable incidents	37265	Dual-chamber implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A071201	Failure to Convert Rhythm	27/05/2025	27/05/2025
									A072201	High Impedance	27/05/2025	27/05/2025
									A072202	Low Impedance	27/05/2025	27/05/2025
									A0908	Incorrect, Inadequate or Imprecise Result or Readings	27/05/2025	27/05/2025
							F08	Hospitalization or Prolonged Hospitalization	A071201	Failure to Convert Rhythm	27/05/2025	27/05/2025
									A072201	High Impedance	27/05/2025	27/05/2025
									A072202	Low Impedance	27/05/2025	27/05/2025
									A0908	Incorrect, Inadequate or Imprecise Result or Readings	27/05/2025	27/05/2025
							F1905	Device Revision or Replacement	A071201	Failure to Convert Rhythm	27/05/2025	27/05/2025
									A072201	High Impedance	27/05/2025	27/05/2025
									A072202	Low Impedance	27/05/2025	27/05/2025
									A0908	Incorrect, Inadequate or Imprecise Result or Readings	27/05/2025	27/05/2025
17057	2025	All other reportable incidents	63188	Left-ventricular wireless pacing system sensor/transmitter	CT145	Cardiovascular devices			A0701	Capturing Problem		
					CT1577	Wireless monitoring system transmitters			A0701	Capturing Problem		
					CT1684	Cardiac pacemakers and associated devices			A0701	Capturing Problem		
					CT2727	Active implant devices/ancillaries			A0701	Capturing Problem		
					CT2750	Implantable cardiac electrical stimulators and associated devices			A0701	Capturing Problem		
					CT2885	Physiological monitoring/measuring system transmitters/receivers/transceivers			A0701	Capturing Problem		
17058	2025	All other reportable incidents	47270	Cardiac resynchronization therapy implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E210401	Shock from Patient Lead(s)	A071302	Inappropriate/Inadequate Shock/Stimulation		
							F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A071302	Inappropriate/Inadequate Shock/Stimulation		
17059	2025	All other reportable incidents	47265	Dual-chamber implantable pacemaker, rate-responsive	CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	18/02/2025	18/02/2025
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	18/02/2025	18/02/2025
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	18/02/2025	18/02/2025
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	18/02/2025	18/02/2025
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	18/02/2025	18/02/2025
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	18/02/2025	18/02/2025
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	18/02/2025	18/02/2025
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	18/02/2025	18/02/2025
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	18/02/2025	18/02/2025
17060	2025		37265	Dual-chamber implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
17061	2025	All other reportable incidents	16995	Intracardiac electrode	CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A1802	Device Contaminated During Manufacture or Shipping		
							F26	No Health Consequences or Impact	A1802	Device Contaminated During Manufacture or Shipping		
17062	2025	All other reportable incidents	37265	Dual-chamber implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A1605	Reset Problem		
									A160503	Inappropriate or Unexpected Reset		
							F26	No Health Consequences or Impact	A1605	Reset Problem		
									A160503	Inappropriate or Unexpected Reset		
17063	2025	All other reportable incidents	47270	Cardiac resynchronization therapy implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E210401	Shock from Patient Lead(s)	A070909	Oversensing		
									A071302	Inappropriate/Inadequate Shock/Stimulation		
									A072201	High Impedance		
									A0908	Incorrect, Inadequate or Imprecise Result or Readings		
							F08	Hospitalization or Prolonged Hospitalization	A070909	Oversensing		
									A071302	Inappropriate/Inadequate Shock/Stimulation		
									A072201	High Impedance		
									A0908	Incorrect, Inadequate or Imprecise Result or Readings		
							F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A070909	Oversensing		
									A071302	Inappropriate/Inadequate Shock/Stimulation		
									A072201	High Impedance		
									A0908	Incorrect, Inadequate or Imprecise Result or Readings		
							F1905	Device Revision or Replacement	A070909	Oversensing		
									A071302	Inappropriate/Inadequate Shock/Stimulation		
									A072201	High Impedance		
									A0908	Incorrect, Inadequate or Imprecise Result or Readings		
							F23	Unexpected Medical Intervention	A070909	Oversensing		
									A071302	Inappropriate/Inadequate Shock/Stimulation		
									A072201	High Impedance		
									A0908	Incorrect, Inadequate or Imprecise Result or Readings		
17064	2025	All other reportable incidents	35852	Single-chamber implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A072201	High Impedance		
							F26	No Health Consequences or Impact	A072201	High Impedance		
17065	2025	All other reportable incidents	47265	Dual-chamber implantable pacemaker, rate-	CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A05	Mechanical Problem	05/06/2025	05/06/2025

				responsive			F26	No Health Consequences or Impact	A05	Mechanical Problem	05/06/2025	05/06/2025
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A05	Mechanical Problem	05/06/2025	05/06/2025
							F26	No Health Consequences or Impact	A05	Mechanical Problem	05/06/2025	05/06/2025
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A05	Mechanical Problem	05/06/2025	05/06/2025
							F26	No Health Consequences or Impact	A05	Mechanical Problem	05/06/2025	05/06/2025
17066	2025	All other reportable incidents	47265	Dual-chamber implantable pacemaker, rate-responsive	CT1685	Cardiac pacemakers	E0606	Cardiomyopathy	A0712	Pacing Problem		
							F12	Serious Injury/ Illness/ Impairment	A0712	Pacing Problem		
							F2307	Reprogramming of Device	A0712	Pacing Problem		
					CT257	Implantable cardiac pacemakers	E0606	Cardiomyopathy	A0712	Pacing Problem		
							F12	Serious Injury/ Illness/ Impairment	A0712	Pacing Problem		
							F2307	Reprogramming of Device	A0712	Pacing Problem		
					CT2866	Implantable cardiac electrical stimulators	E0606	Cardiomyopathy	A0712	Pacing Problem		
							F12	Serious Injury/ Illness/ Impairment	A0712	Pacing Problem		
							F2307	Reprogramming of Device	A0712	Pacing Problem		
17067	2025	All other reportable incidents	47267	Single-chamber implantable pacemaker, rate-responsive	CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem	02/05/2025	02/05/2025
									A070504	Premature Discharge of Battery	02/05/2025	02/05/2025
							F08	Hospitalization or Prolonged Hospitalization	A0705	Battery Problem	02/05/2025	02/05/2025
									A070504	Premature Discharge of Battery	02/05/2025	02/05/2025
							F1905	Device Revision or Replacement	A0705	Battery Problem	02/05/2025	02/05/2025
									A070504	Premature Discharge of Battery	02/05/2025	02/05/2025
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem	02/05/2025	02/05/2025
									A070504	Premature Discharge of Battery	02/05/2025	02/05/2025
							F08	Hospitalization or Prolonged Hospitalization	A0705	Battery Problem	02/05/2025	02/05/2025
									A070504	Premature Discharge of Battery	02/05/2025	02/05/2025
							F1905	Device Revision or Replacement	A0705	Battery Problem	02/05/2025	02/05/2025
									A070504	Premature Discharge of Battery	02/05/2025	02/05/2025
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem	02/05/2025	02/05/2025
									A070504	Premature Discharge of Battery	02/05/2025	02/05/2025
							F08	Hospitalization or Prolonged Hospitalization	A0705	Battery Problem	02/05/2025	02/05/2025
									A070504	Premature Discharge of Battery	02/05/2025	02/05/2025
							F1905	Device Revision or Replacement	A0705	Battery Problem	02/05/2025	02/05/2025
									A070504	Premature Discharge of Battery	02/05/2025	02/05/2025
17068	2025	All other reportable incidents	37265	Dual-chamber implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A0709	Device Sensing Problem		
									A071302	Inappropriate/Inadequate Shock/Stimulation		
							F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A0709	Device Sensing Problem		
									A071302	Inappropriate/Inadequate Shock/Stimulation		
							F1203	Life Threatening Illness or Injury	A0709	Device Sensing Problem		
									A071302	Inappropriate/Inadequate Shock/Stimulation		
							F2307	Reprogramming of Device	A0709	Device Sensing Problem		
									A071302	Inappropriate/Inadequate Shock/Stimulation		
17069	2025	All other reportable incidents	47270	Cardiac resynchronization therapy implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
									A0908	Incorrect, Inadequate or Imprecise Result or Readings		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
									A0908	Incorrect, Inadequate or Imprecise Result or Readings		
17070	2025	All other reportable incidents	47265	Dual-chamber implantable pacemaker, rate-responsive	CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A0908	Incorrect, Inadequate or Imprecise Result or Readings		
							F26	No Health Consequences or Impact	A0908	Incorrect, Inadequate or Imprecise Result or Readings		
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A0908	Incorrect, Inadequate or Imprecise Result or Readings		
							F26	No Health Consequences or Impact	A0908	Incorrect, Inadequate or Imprecise Result or Readings		
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A0908	Incorrect, Inadequate or Imprecise Result or Readings		
							F26	No Health Consequences or Impact	A0908	Incorrect, Inadequate or Imprecise Result or Readings		
17071	2025	All other reportable incidents	47265	Dual-chamber implantable pacemaker, rate-responsive			E2403	No Clinical Signs, Symptoms or Conditions	A05	Mechanical Problem	04/06/2025	04/06/2025
									A070101	Failure to Capture	04/06/2025	04/06/2025
									A0712	Pacing Problem	04/06/2025	04/06/2025
							F26	No Health Consequences or Impact	A05	Mechanical Problem	04/06/2025	04/06/2025
									A070101	Failure to Capture	04/06/2025	04/06/2025
									A0712	Pacing Problem	04/06/2025	04/06/2025
17072	2025	All other reportable incidents	37265	Dual-chamber implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A072201	High Impedance		
									A0908	Incorrect, Inadequate or Imprecise Result or Readings		
							F26	No Health Consequences or Impact	A072201	High Impedance		
									A0908	Incorrect, Inadequate or Imprecise Result or Readings		
17073	2025	All other reportable incidents	47265	Dual-chamber implantable pacemaker, rate-responsive	CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A13	Communication or Transmission Problem	05/06/2025	05/06/2025
							F26	No Health Consequences or Impact	A13	Communication or Transmission Problem	05/06/2025	05/06/2025
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A13	Communication or Transmission Problem	05/06/2025	05/06/2025
							F26	No Health Consequences or Impact	A13	Communication or Transmission Problem	05/06/2025	05/06/2025
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A13	Communication or Transmission Problem	05/06/2025	05/06/2025
							F26	No Health Consequences or Impact	A13	Communication or Transmission Problem	05/06/2025	05/06/2025
17074	2025	All other reportable incidents	47263	Cardiac resynchronization therapy implantable pacemaker	CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A0908	Incorrect, Inadequate or Imprecise Result or Readings	14/05/2024	14/05/2024
							F08	Hospitalization or Prolonged Hospitalization	A0908	Incorrect, Inadequate or Imprecise Result or Readings	14/05/2024	14/05/2024
							F1905	Device Revision or Replacement	A0908	Incorrect, Inadequate or Imprecise Result or Readings	14/05/2024	14/05/2024
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A0908	Incorrect, Inadequate or Imprecise Result or Readings	14/05/2024	14/05/2024
							F08	Hospitalization or Prolonged Hospitalization	A0908	Incorrect, Inadequate or Imprecise Result or Readings	14/05/2024	14/05/2024
							F1905	Device Revision or Replacement	A0908	Incorrect, Inadequate or Imprecise Result or Readings	14/05/2024	14/05/2024
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A0908	Incorrect, Inadequate or Imprecise Result or Readings	14/05/2024	14/05/2024
							F08	Hospitalization or Prolonged Hospitalization	A0908	Incorrect, Inadequate or Imprecise Result or Readings	14/05/2024	14/05/2024

							F1905	Device Revision or Replacement	A0908	Incorrect, Inadequate or Imprecise Result or Readings	14/05/2024	14/05/2024
17075	2025	All other reportable incidents	47270	Cardiac resynchronization therapy implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
									A0908	Incorrect, Inadequate or Imprecise Result or Readings		
							F26	No Health Consequences or Impact	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
									A0908	Incorrect, Inadequate or Imprecise Result or Readings		
17076	2025	All other reportable incidents	47270	Cardiac resynchronization therapy implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	30/05/2025	30/05/2025
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	30/05/2025	30/05/2025
							F1903	Device Explantation	A070504	Premature Discharge of Battery	30/05/2025	30/05/2025
17077	2025	All other reportable incidents	37265	Dual-chamber implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	13/08/2025	13/08/2025
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	13/08/2025	13/08/2025
							F1903	Device Explantation	A070504	Premature Discharge of Battery	13/08/2025	13/08/2025
17078	2025		18145	Cardiac myoplasty electrical stimulation system	CT1684	Cardiac pacemakers and associated devices	F26	No Health Consequences or Impact	A0701	Capturing Problem		
									A071207	Pacing Inadequately		
					CT2866	Implantable cardiac electrical stimulators	F26	No Health Consequences or Impact	A0701	Capturing Problem		
									A071207	Pacing Inadequately		
17079	2025	All other reportable incidents	17882	Physiologic-monitoring defibrillation system	CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070803	Failure to Power Up		
							F27	Problem Identified During Non-Clinical Procedure	A070803	Failure to Power Up		
17080	2025	All other reportable incidents	47265	Dual-chamber implantable pacemaker, rate-responsive			E060104	Bradycardia	A160601	Premature Elective Replacement Indicator	30/05/2025	30/05/2025
							F1905	Device Revision or Replacement	A160601	Premature Elective Replacement Indicator	30/05/2025	30/05/2025
17081	2025	All other reportable incidents	37265	Dual-chamber implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A070909	Oversensing		
									A072001	Electromagnetic Interference		
							F12	Serious Injury/ Illness/ Impairment	A070909	Oversensing		
									A072001	Electromagnetic Interference		
							F2307	Reprogramming of Device	A070909	Oversensing		
									A072001	Electromagnetic Interference		
							F26	No Health Consequences or Impact	A070909	Oversensing		
									A072001	Electromagnetic Interference		
17082	2025	All other reportable incidents	47270	Cardiac resynchronization therapy implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
									A0908	Incorrect, Inadequate or Imprecise Result or Readings		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
									A0908	Incorrect, Inadequate or Imprecise Result or Readings		
17083	2025	All other reportable incidents	47270	Cardiac resynchronization therapy implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A070102	High Capture Threshold		
									A070909	Oversensing		
									A070910	Undersensing		
							F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A070102	High Capture Threshold		
									A070909	Oversensing		
									A070910	Undersensing		
17084	2025	All other reportable incidents	47270	Cardiac resynchronization therapy implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A1304	Telemetry Discrepancy		
							F26	No Health Consequences or Impact	A1304	Telemetry Discrepancy		
17085	2025	All other reportable incidents	17882	Physiologic-monitoring defibrillation system	CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A1205	Loose or Intermittent Connection		
							F27	Problem Identified During Non-Clinical Procedure	A1205	Loose or Intermittent Connection		
17086	2025		60789	Intracardiac pacemaker	CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem		
							F26	No Health Consequences or Impact	A0705	Battery Problem		
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem		
							F26	No Health Consequences or Impact	A0705	Battery Problem		
					CT2818	Catheterization/Cannulation/Fluid path devices	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem		
							F26	No Health Consequences or Impact	A0705	Battery Problem		
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem		
							F26	No Health Consequences or Impact	A0705	Battery Problem		
17087	2025	All other reportable incidents	36106	Implantable pacemaker analyser, intraoperative	CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A09	Output Problem		
									A1104	Computer Operating System Problem		
									A13	Communication or Transmission Problem		
							F26	No Health Consequences or Impact	A09	Output Problem		
									A1104	Computer Operating System Problem		
									A13	Communication or Transmission Problem		
17088	2025	All other reportable incidents	47205	Cardiac pulse generator programmer	CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A13	Communication or Transmission Problem		
							F26	No Health Consequences or Impact	A13	Communication or Transmission Problem		
17089	2025	All other reportable incidents	47270	Cardiac resynchronization therapy implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A1107	Data Problem		
							F26	No Health Consequences or Impact	A1107	Data Problem		
17090	2025	All other reportable incidents	47270	Cardiac resynchronization therapy implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A0709	Device Sensing Problem		
									A0712	Pacing Problem		
							F12	Serious Injury/ Illness/ Impairment	A0709	Device Sensing Problem		
									A0712	Pacing Problem		
							F2307	Reprogramming of Device	A0709	Device Sensing Problem		
									A0712	Pacing Problem		
							F26	No Health Consequences or Impact	A0709	Device Sensing Problem		
									A0712	Pacing Problem		
17091	2025	All other reportable incidents	47270	Cardiac resynchronization therapy implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E210401	Shock from Patient Lead(s)	A070909	Oversensing		
									A071302	Inappropriate/Inadequate Shock/Stimulation		
									A0909	Unexpected Therapeutic Results		
							F08	Hospitalization or Prolonged Hospitalization	A070909	Oversensing		
									A071302	Inappropriate/Inadequate Shock/Stimulation		
									A0909	Unexpected Therapeutic Results		
							F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A070909	Oversensing		
									A071302	Inappropriate/Inadequate Shock/Stimulation		
									A0909	Unexpected Therapeutic Results		
							F2303	Medication Required	A070909	Oversensing		
									A071302	Inappropriate/Inadequate Shock/Stimulation		
									A0909	Unexpected Therapeutic Results		
17092	2025	All other reportable incidents	47265	Dual-chamber implantable pacemaker, rate-responsive			E0112	Dizziness	A0908	Incorrect, Inadequate or Imprecise Result or Readings	06/06/2025	06/06/2025

							F08	Hospitalization or Prolonged Hospitalization	A0908	Incorrect, Inadequate or Imprecise Result or Readings	06/06/2025	06/06/2025
							F1905	Device Revision or Replacement	A0908	Incorrect, Inadequate or Imprecise Result or Readings	06/06/2025	06/06/2025
17093	2025	All other reportable incidents	47270	Cardiac resynchronization therapy implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E210401	Shock from Patient Lead(s)	A070909	Oversensing		
									A071302	Inappropriate/Inadequate Shock/Stimulation		
									A0909	Unexpected Therapeutic Results		
							F08	Hospitalization or Prolonged Hospitalization	A070909	Oversensing		
									A071302	Inappropriate/Inadequate Shock/Stimulation		
									A0909	Unexpected Therapeutic Results		
							F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A070909	Oversensing		
									A071302	Inappropriate/Inadequate Shock/Stimulation		
									A0909	Unexpected Therapeutic Results		
							F2303	Medication Required	A070909	Oversensing		
									A071302	Inappropriate/Inadequate Shock/Stimulation		
									A0909	Unexpected Therapeutic Results		
17094	2025	All other reportable incidents	47270	Cardiac resynchronization therapy implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A051201	Device Dislodged or Dislocated		
							F26	No Health Consequences or Impact	A051201	Device Dislodged or Dislocated		
17095	2025	All other reportable incidents	47270	Cardiac resynchronization therapy implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A010402	Migration		
									A0908	Incorrect, Inadequate or Imprecise Result or Readings		
									A1107	Data Problem		
							F26	No Health Consequences or Impact	A010402	Migration		
									A0908	Incorrect, Inadequate or Imprecise Result or Readings		
									A1107	Data Problem		
17096	2025	All other reportable incidents	37265	Dual-chamber implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E210401	Shock from Patient Lead(s)	A070909	Oversensing		
									A0712	Pacing Problem		
									A071302	Inappropriate/Inadequate Shock/Stimulation		
							F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A070909	Oversensing		
									A0712	Pacing Problem		
									A071302	Inappropriate/Inadequate Shock/Stimulation		
17097	2025	All other reportable incidents	37265	Dual-chamber implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
17098	2025	All other reportable incidents	35852	Single-chamber implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A070909	Oversensing	04/06/2025	04/06/2025
							F08	Hospitalization or Prolonged Hospitalization	A070909	Oversensing	04/06/2025	04/06/2025
							F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A070909	Oversensing	04/06/2025	04/06/2025
							F1905	Device Revision or Replacement	A070909	Oversensing	04/06/2025	04/06/2025
							F2203	Imaging Required	A070909	Oversensing	04/06/2025	04/06/2025
17099	2025	All other reportable incidents	47270	Cardiac resynchronization therapy implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E210401	Shock from Patient Lead(s)	A071302	Inappropriate/Inadequate Shock/Stimulation		
							F08	Hospitalization or Prolonged Hospitalization	A071302	Inappropriate/Inadequate Shock/Stimulation		
							F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A071302	Inappropriate/Inadequate Shock/Stimulation		
17100	2025	All other reportable incidents	47270	Cardiac resynchronization therapy implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A0712	Pacing Problem		
									A072001	Electromagnetic Interference		
							F26	No Health Consequences or Impact	A0712	Pacing Problem		
									A072001	Electromagnetic Interference		
17101	2025		47265	Dual-chamber implantable pacemaker, rate-responsive	CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A071102	Failure to Interrogate		
							F22	Unexpected Diagnostic Intervention	A071102	Failure to Interrogate		
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A071102	Failure to Interrogate		
							F22	Unexpected Diagnostic Intervention	A071102	Failure to Interrogate		
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A071102	Failure to Interrogate		
							F22	Unexpected Diagnostic Intervention	A071102	Failure to Interrogate		
17102	2025	All other reportable incidents	47270	Cardiac resynchronization therapy implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A072201	High Impedance		
							F26	No Health Consequences or Impact	A072201	High Impedance		
17103	2025	All other reportable incidents	47265	Dual-chamber implantable pacemaker, rate-responsive	CT1685	Cardiac pacemakers	E060104	Bradycardia	A0908	Incorrect, Inadequate or Imprecise Result or Readings	13/06/2025	13/06/2025
							E0717	Dyspnea	A0908	Incorrect, Inadequate or Imprecise Result or Readings	13/06/2025	13/06/2025
							F08	Hospitalization or Prolonged Hospitalization	A0908	Incorrect, Inadequate or Imprecise Result or Readings	13/06/2025	13/06/2025
							F1905	Device Revision or Replacement	A0908	Incorrect, Inadequate or Imprecise Result or Readings	13/06/2025	13/06/2025
					CT257	Implantable cardiac pacemakers	E060104	Bradycardia	A0908	Incorrect, Inadequate or Imprecise Result or Readings	13/06/2025	13/06/2025
							E0717	Dyspnea	A0908	Incorrect, Inadequate or Imprecise Result or Readings	13/06/2025	13/06/2025
							F08	Hospitalization or Prolonged Hospitalization	A0908	Incorrect, Inadequate or Imprecise Result or Readings	13/06/2025	13/06/2025
							F1905	Device Revision or Replacement	A0908	Incorrect, Inadequate or Imprecise Result or Readings	13/06/2025	13/06/2025
					CT2866	Implantable cardiac electrical stimulators	E060104	Bradycardia	A0908	Incorrect, Inadequate or Imprecise Result or Readings	13/06/2025	13/06/2025
							E0717	Dyspnea	A0908	Incorrect, Inadequate or Imprecise Result or Readings	13/06/2025	13/06/2025
							F08	Hospitalization or Prolonged Hospitalization	A0908	Incorrect, Inadequate or Imprecise Result or Readings	13/06/2025	13/06/2025
							F1905	Device Revision or Replacement	A0908	Incorrect, Inadequate or Imprecise Result or Readings	13/06/2025	13/06/2025
17104	2025	All other reportable incidents	35852	Single-chamber implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A072202	Low Impedance		
							F2307	Reprogramming of Device	A072202	Low Impedance		
17105	2025	Unanticipated serious deterioration in state of health	47265	Dual-chamber implantable pacemaker, rate-responsive	CT1685	Cardiac pacemakers	E0505	Hematoma	A24	No Apparent Device Problem	12/06/2025	12/06/2025
							E1718	Skin Erosion	A24	No Apparent Device Problem	12/06/2025	12/06/2025
							F08	Hospitalization or Prolonged Hospitalization	A24	No Apparent Device Problem	12/06/2025	12/06/2025
							F12	Serious Injury/ Illness/ Impairment	A24	No Apparent Device Problem	12/06/2025	12/06/2025
							F1903	Device Explantation	A24	No Apparent Device Problem	12/06/2025	12/06/2025
							F26	No Health Consequences or Impact	A24	No Apparent Device Problem	12/06/2025	12/06/2025
					CT257	Implantable cardiac pacemakers	E0505	Hematoma	A24	No Apparent Device Problem	12/06/2025	12/06/2025
							E1718	Skin Erosion	A24	No Apparent Device Problem	12/06/2025	12/06/2025
							F08	Hospitalization or Prolonged Hospitalization	A24	No Apparent Device Problem	12/06/2025	12/06/2025
							F12	Serious Injury/ Illness/ Impairment	A24	No Apparent Device Problem	12/06/2025	12/06/2025

							F1903	Device Explantation	A24	No Apparent Device Problem	12/06/2025	12/06/2025
							F26	No Health Consequences or Impact	A24	No Apparent Device Problem	12/06/2025	12/06/2025
					CT2866	Implantable cardiac electrical stimulators	E0505	Hematoma	A24	No Apparent Device Problem	12/06/2025	12/06/2025
							E1718	Skin Erosion	A24	No Apparent Device Problem	12/06/2025	12/06/2025
							F08	Hospitalization or Prolonged Hospitalization	A24	No Apparent Device Problem	12/06/2025	12/06/2025
							F12	Serious Injury/ Illness/ Impairment	A24	No Apparent Device Problem	12/06/2025	12/06/2025
							F1903	Device Explantation	A24	No Apparent Device Problem	12/06/2025	12/06/2025
							F26	No Health Consequences or Impact	A24	No Apparent Device Problem	12/06/2025	12/06/2025
17106	2025	All other reportable incidents	47205	Cardiac pulse generator programmer	CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A13	Communication or Transmission Problem		
							F26	No Health Consequences or Impact	A13	Communication or Transmission Problem		
17107	2025	All other reportable incidents	35852	Single-chamber implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A072001	Electromagnetic Interference		
									A090801	Signal Artifact/Noise		
							F26	No Health Consequences or Impact	A072001	Electromagnetic Interference		
									A090801	Signal Artifact/Noise		
17108	2025	All other reportable incidents	60789	Intracardiac pacemaker	CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	19/06/2025	19/06/2025
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	19/06/2025	19/06/2025
							F12	Serious Injury/ Illness/ Impairment	A070504	Premature Discharge of Battery	19/06/2025	19/06/2025
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	19/06/2025	19/06/2025
							F23	Unexpected Medical Intervention	A070504	Premature Discharge of Battery	19/06/2025	19/06/2025
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	19/06/2025	19/06/2025
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	19/06/2025	19/06/2025
							F12	Serious Injury/ Illness/ Impairment	A070504	Premature Discharge of Battery	19/06/2025	19/06/2025
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	19/06/2025	19/06/2025
							F23	Unexpected Medical Intervention	A070504	Premature Discharge of Battery	19/06/2025	19/06/2025
					CT2818	Catheterization/Cannulation/Fluid path devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	19/06/2025	19/06/2025
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	19/06/2025	19/06/2025
							F12	Serious Injury/ Illness/ Impairment	A070504	Premature Discharge of Battery	19/06/2025	19/06/2025
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	19/06/2025	19/06/2025
							F23	Unexpected Medical Intervention	A070504	Premature Discharge of Battery	19/06/2025	19/06/2025
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	19/06/2025	19/06/2025
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	19/06/2025	19/06/2025
							F12	Serious Injury/ Illness/ Impairment	A070504	Premature Discharge of Battery	19/06/2025	19/06/2025
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	19/06/2025	19/06/2025
							F23	Unexpected Medical Intervention	A070504	Premature Discharge of Battery	19/06/2025	19/06/2025
17109	2025	All other reportable incidents	37265	Dual-chamber implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A070909	Oversensing		
									A072001	Electromagnetic Interference		
							F12	Serious Injury/ Illness/ Impairment	A070909	Oversensing		
									A072001	Electromagnetic Interference		
							F2307	Reprogramming of Device	A070909	Oversensing		
									A072001	Electromagnetic Interference		
17110	2025	All other reportable incidents	47265	Dual-chamber implantable pacemaker, rate-responsive			E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	06/06/2025	06/06/2025
									A0908	Incorrect, Inadequate or Imprecise Result or Readings	06/06/2025	06/06/2025
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	06/06/2025	06/06/2025
									A0908	Incorrect, Inadequate or Imprecise Result or Readings	06/06/2025	06/06/2025
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	06/06/2025	06/06/2025
									A0908	Incorrect, Inadequate or Imprecise Result or Readings	06/06/2025	06/06/2025
17111	2025		47265	Dual-chamber implantable pacemaker, rate-responsive	CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	11/06/2025	11/06/2025
							F1901	Additional Surgery	A24	No Apparent Device Problem	11/06/2025	11/06/2025
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	11/06/2025	11/06/2025
							F1901	Additional Surgery	A24	No Apparent Device Problem	11/06/2025	11/06/2025
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	11/06/2025	11/06/2025
							F1901	Additional Surgery	A24	No Apparent Device Problem	11/06/2025	11/06/2025
17112	2025	All other reportable incidents	47265	Dual-chamber implantable pacemaker, rate-responsive			E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	01/06/2025	01/06/2025
									A0908	Incorrect, Inadequate or Imprecise Result or Readings	01/06/2025	01/06/2025
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	01/06/2025	01/06/2025
									A0908	Incorrect, Inadequate or Imprecise Result or Readings	01/06/2025	01/06/2025
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	01/06/2025	01/06/2025
									A0908	Incorrect, Inadequate or Imprecise Result or Readings	01/06/2025	01/06/2025
17113	2025	Unanticipated serious deterioration in state of health	46554	Pacing/defibrillation lead repair tool	CT1684	Cardiac pacemakers and associated devices	E0602	Cardiac Arrest	A070504	Premature Discharge of Battery	04/06/2025	04/06/2025
							F02	Death	A070504	Premature Discharge of Battery	04/06/2025	04/06/2025
17114	2025	All other reportable incidents	47270	Cardiac resynchronization therapy implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A070909	Oversensing		
									A071204	No Pacing		
							F23	Unexpected Medical Intervention	A070909	Oversensing		
									A071204	No Pacing		
17115	2025	All other reportable incidents	47270	Cardiac resynchronization therapy implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E210401	Shock from Patient Lead(s)	A0709	Device Sensing Problem		
									A0712	Pacing Problem		
									A071302	Inappropriate/Inadequate Shock/Stimulation		
							F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A0709	Device Sensing Problem		
									A0712	Pacing Problem		
									A071302	Inappropriate/Inadequate Shock/Stimulation		
							F1203	Life Threatening Illness or Injury	A0709	Device Sensing Problem		
									A0712	Pacing Problem		
									A071302	Inappropriate/Inadequate Shock/Stimulation		
17116	2025	All other reportable incidents	47265	Dual-chamber implantable pacemaker, rate-responsive	CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A13	Communication or Transmission Problem		
							F26	No Health Consequences or Impact	A13	Communication or Transmission Problem		
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A13	Communication or Transmission Problem		
							F26	No Health Consequences or Impact	A13	Communication or Transmission Problem		
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A13	Communication or Transmission Problem		
							F26	No Health Consequences or Impact	A13	Communication or Transmission Problem		
17117	2025	All other reportable incidents	47270	Cardiac resynchronization therapy implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A072201	High Impedance		
							F2203	Imaging Required	A072201	High Impedance		
17118	2025	All other reportable incidents	60789	Intracardiac pacemaker	CT1685	Cardiac pacemakers	E0619	Pericardial Effusion	A090811	Unable to Obtain Readings		
							F08	Hospitalization or Prolonged Hospitalization	A090811	Unable to Obtain Readings		

					CT257	Implantable cardiac pacemakers	E0619	Pericardial Effusion	A090811	Unable to Obtain Readings		
							F08	Hospitalization or Prolonged Hospitalization	A090811	Unable to Obtain Readings		
					CT2818	Catheterization/Cannulation/Fluid path devices	E0619	Pericardial Effusion	A090811	Unable to Obtain Readings		
							F08	Hospitalization or Prolonged Hospitalization	A090811	Unable to Obtain Readings		
					CT2866	Implantable cardiac electrical stimulators	E0619	Pericardial Effusion	A090811	Unable to Obtain Readings		
							F08	Hospitalization or Prolonged Hospitalization	A090811	Unable to Obtain Readings		
17119	2025	All other reportable incidents	16995	Intracardiac electrode	CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A1802	Device Contaminated During Manufacture or Shipping		
							F26	No Health Consequences or Impact	A1802	Device Contaminated During Manufacture or Shipping		
17120	2025	All other reportable incidents	17882	Physiologic-monitoring defibrillation system	CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070912	Incorrect Interpretation of Signal		
							F26	No Health Consequences or Impact	A070912	Incorrect Interpretation of Signal		
17121	2025	All other reportable incidents	47205	Cardiac pulse generator programmer	CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A150105	Key or Button Unresponsive/not Working		
							F26	No Health Consequences or Impact	A150105	Key or Button Unresponsive/not Working		
17122	2025	All other reportable incidents	47270	Cardiac resynchronization therapy implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A05	Mechanical Problem		
							F26	No Health Consequences or Impact	A05	Mechanical Problem		
17123	2025	All other reportable incidents	47270	Cardiac resynchronization therapy implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A0709	Device Sensing Problem		
							F26	No Health Consequences or Impact	A0709	Device Sensing Problem		
17124	2025	All other reportable incidents	47265	Dual-chamber implantable pacemaker, rate-responsive	CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	03/06/2025	03/06/2025
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	03/06/2025	03/06/2025
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	03/06/2025	03/06/2025
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	03/06/2025	03/06/2025
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	03/06/2025	03/06/2025
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	03/06/2025	03/06/2025
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	03/06/2025	03/06/2025
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	03/06/2025	03/06/2025
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	03/06/2025	03/06/2025
17125	2025	All other reportable incidents	35852	Single-chamber implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem		
									A0908	Incorrect, Inadequate or Imprecise Result or Readings		
							F26	No Health Consequences or Impact	A0705	Battery Problem		
									A0908	Incorrect, Inadequate or Imprecise Result or Readings		
17126	2025	All other reportable incidents	47265	Dual-chamber implantable pacemaker, rate-responsive	CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A0104	Migration or Expulsion of Device		
									A1605	Reset Problem		
							F2307	Reprogramming of Device	A0104	Migration or Expulsion of Device		
									A1605	Reset Problem		
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A0104	Migration or Expulsion of Device		
									A1605	Reset Problem		
							F2307	Reprogramming of Device	A0104	Migration or Expulsion of Device		
									A1605	Reset Problem		
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A0104	Migration or Expulsion of Device		
									A1605	Reset Problem		
							F2307	Reprogramming of Device	A0104	Migration or Expulsion of Device		
									A1605	Reset Problem		
17127	2025	All other reportable incidents	17882	Physiologic-monitoring defibrillation system	CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070912	Incorrect Interpretation of Signal		
							F26	No Health Consequences or Impact	A070912	Incorrect Interpretation of Signal		
17128	2025	All other reportable incidents	47265	Dual-chamber implantable pacemaker, rate-responsive			E2103	Device Overstimulation of Tissue	A070504	Premature Discharge of Battery	20/06/2025	20/06/2025
									A0712	Pacing Problem	20/06/2025	20/06/2025
									A0908	Incorrect, Inadequate or Imprecise Result or Readings	20/06/2025	20/06/2025
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	20/06/2025	20/06/2025
									A0712	Pacing Problem	20/06/2025	20/06/2025
									A0908	Incorrect, Inadequate or Imprecise Result or Readings	20/06/2025	20/06/2025
							F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A070504	Premature Discharge of Battery	20/06/2025	20/06/2025
									A0712	Pacing Problem	20/06/2025	20/06/2025
									A0908	Incorrect, Inadequate or Imprecise Result or Readings	20/06/2025	20/06/2025
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	20/06/2025	20/06/2025
									A0712	Pacing Problem	20/06/2025	20/06/2025
									A0908	Incorrect, Inadequate or Imprecise Result or Readings	20/06/2025	20/06/2025
17129	2025	All other reportable incidents	37265	Dual-chamber implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A072001	Electromagnetic Interference		
							F26	No Health Consequences or Impact	A072001	Electromagnetic Interference		
17130	2025	All other reportable incidents	47205	Cardiac pulse generator programmer	CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A072201	High Impedance		
							F2203	Imaging Required	A072201	High Impedance		
17131	2025	All other reportable incidents	36106	Implantable pacemaker analyser, intraoperative	CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A1104	Computer Operating System Problem		
							F26	No Health Consequences or Impact	A1104	Computer Operating System Problem		
17132	2025		17882	Physiologic-monitoring defibrillation system	CT1684	Cardiac pacemakers and associated devices	F12	Serious Injury/ Illness/ Impairment	A070912	Incorrect Interpretation of Signal		
									A12	Connection Problem		
17133	2025	All other reportable incidents	47265	Dual-chamber implantable pacemaker, rate-responsive	CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A070909	Oversensing		
							F26	No Health Consequences or Impact	A070909	Oversensing		
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A070909	Oversensing		
							F26	No Health Consequences or Impact	A070909	Oversensing		
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A070909	Oversensing		
							F26	No Health Consequences or Impact	A070909	Oversensing		
17134	2025	All other reportable incidents	47265	Dual-chamber implantable pacemaker, rate-responsive	CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem		
							F1901	Additional Surgery	A24	No Apparent Device Problem		
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem		
							F1901	Additional Surgery	A24	No Apparent Device Problem		
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem		
							F1901	Additional Surgery	A24	No Apparent Device Problem		
17135	2025	All other reportable incidents	61126	Implantable pulse generator mesh bag, bioabsorbable	CT1684	Cardiac pacemakers and associated devices	E1906	Unspecified Infection	A24	No Apparent Device Problem		
							E2338	Swelling/ Edema	A24	No Apparent Device Problem		
							F12	Serious Injury/ Illness/ Impairment	A24	No Apparent Device Problem		
							F2303	Medication Required	A24	No Apparent Device Problem		
17136	2025	All other reportable incidents	61126	Implantable pulse generator mesh bag, bioabsorbable	CT1684	Cardiac pacemakers and associated devices	E1906	Unspecified Infection	A24	No Apparent Device Problem		
							E2338	Swelling/ Edema	A24	No Apparent Device Problem		
							F12	Serious Injury/ Illness/ Impairment	A24	No Apparent Device Problem		

							F1903	Device Explantation	A24	No Apparent Device Problem		
							F2303	Medication Required	A24	No Apparent Device Problem		
17137	2025	All other reportable incidents	61126	Implantable pulse generator mesh bag, bioabsorbable	CT1684	Cardiac pacemakers and associated devices	E1906	Unspecified Infection	A24	No Apparent Device Problem		
							E2338	Swelling/ Edema	A24	No Apparent Device Problem		
							F12	Serious Injury/ Illness/ Impairment	A24	No Apparent Device Problem		
							F2303	Medication Required	A24	No Apparent Device Problem		
17138	2025	All other reportable incidents	61126	Implantable pulse generator mesh bag, bioabsorbable	CT1684	Cardiac pacemakers and associated devices	E1906	Unspecified Infection	A24	No Apparent Device Problem		
							E2338	Swelling/ Edema	A24	No Apparent Device Problem		
							F12	Serious Injury/ Illness/ Impairment	A24	No Apparent Device Problem		
							F2303	Medication Required	A24	No Apparent Device Problem		
17139	2025	All other reportable incidents	61126	Implantable pulse generator mesh bag, bioabsorbable	CT1684	Cardiac pacemakers and associated devices	E1906	Unspecified Infection	A24	No Apparent Device Problem		
							E2338	Swelling/ Edema	A24	No Apparent Device Problem		
							F12	Serious Injury/ Illness/ Impairment	A24	No Apparent Device Problem		
							F1903	Device Explantation	A24	No Apparent Device Problem		
							F2303	Medication Required	A24	No Apparent Device Problem		
17140	2025	All other reportable incidents	17882	Physiologic-monitoring defibrillation system	CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A0712	Pacing Problem		
							F27	Problem Identified During Non-Clinical Procedure	A0712	Pacing Problem		
17141	2025	All other reportable incidents	35852	Single-chamber implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E210401	Shock from Patient Lead(s)	A072201	High Impedance		
							F17	Sedation	A072201	High Impedance		
							F22	Unexpected Diagnostic Intervention	A072201	High Impedance		
17142	2025	All other reportable incidents	35852	Single-chamber implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
17143	2025	All other reportable incidents	47265	Dual-chamber implantable pacemaker, rate-responsive	CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A0203	Defective Device	15/08/2024	15/08/2024
							F24	Insufficient Health Impact Information	A0203	Defective Device	15/08/2024	15/08/2024
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A0203	Defective Device	15/08/2024	15/08/2024
							F24	Insufficient Health Impact Information	A0203	Defective Device	15/08/2024	15/08/2024
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A0203	Defective Device	15/08/2024	15/08/2024
							F24	Insufficient Health Impact Information	A0203	Defective Device	15/08/2024	15/08/2024
17144	2025	All other reportable incidents	35852	Single-chamber implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E210401	Shock from Patient Lead(s)	A070909	Oversensing		
									A071302	Inappropriate/Inadequate Shock/Stimulation		
									A1301	Failure to Read Input Signal		
							F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A070909	Oversensing		
									A071302	Inappropriate/Inadequate Shock/Stimulation		
									A1301	Failure to Read Input Signal		
17145	2025	All other reportable incidents	47265	Dual-chamber implantable pacemaker, rate-responsive	CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	20/06/2025	20/06/2025
									A0712	Pacing Problem	20/06/2025	20/06/2025
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	20/06/2025	20/06/2025
									A0712	Pacing Problem	20/06/2025	20/06/2025
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	20/06/2025	20/06/2025
									A0712	Pacing Problem	20/06/2025	20/06/2025
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	20/06/2025	20/06/2025
									A0712	Pacing Problem	20/06/2025	20/06/2025
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	20/06/2025	20/06/2025
									A0712	Pacing Problem	20/06/2025	20/06/2025
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	20/06/2025	20/06/2025
									A0712	Pacing Problem	20/06/2025	20/06/2025
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	20/06/2025	20/06/2025
									A0712	Pacing Problem	20/06/2025	20/06/2025
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	20/06/2025	20/06/2025
									A0712	Pacing Problem	20/06/2025	20/06/2025
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	20/06/2025	20/06/2025
									A0712	Pacing Problem	20/06/2025	20/06/2025
17146	2025	All other reportable incidents	37265	Dual-chamber implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E210401	Shock from Patient Lead(s)	A0709	Device Sensing Problem		
									A0712	Pacing Problem		
									A071302	Inappropriate/Inadequate Shock/Stimulation		
							F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A0709	Device Sensing Problem		
									A0712	Pacing Problem		
									A071302	Inappropriate/Inadequate Shock/Stimulation		
							F1203	Life Threatening Illness or Injury	A0709	Device Sensing Problem		
									A0712	Pacing Problem		
									A071302	Inappropriate/Inadequate Shock/Stimulation		
							F2307	Reprogramming of Device	A0709	Device Sensing Problem		
									A0712	Pacing Problem		
									A071302	Inappropriate/Inadequate Shock/Stimulation		
17147	2025	All other reportable incidents	17882	Physiologic-monitoring defibrillation system	CT1684	Cardiac pacemakers and associated devices	E0602	Cardiac Arrest	A090804	False Positive Result		
							F05	Delay to Treatment/ Therapy	A090804	False Positive Result		
							F2301	Additional Device Required	A090804	False Positive Result		
17148	2025	All other reportable incidents	47270	Cardiac resynchronization therapy implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E210401	Shock from Patient Lead(s)	A070909	Oversensing		
									A071302	Inappropriate/Inadequate Shock/Stimulation		
									A090801	Signal Artifact/Noise		
							F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A070909	Oversensing		
									A071302	Inappropriate/Inadequate Shock/Stimulation		
									A090801	Signal Artifact/Noise		
							F2203	Imaging Required	A070909	Oversensing		
									A071302	Inappropriate/Inadequate Shock/Stimulation		
									A090801	Signal Artifact/Noise		
17149	2025	All other reportable incidents	35852	Single-chamber implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
17150	2025	All other reportable incidents	47270	Cardiac resynchronization therapy implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E210401	Shock from Patient Lead(s)	A070909	Oversensing		
									A071302	Inappropriate/Inadequate Shock/Stimulation		
									A090801	Signal Artifact/Noise		
							F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A070909	Oversensing		
									A071302	Inappropriate/Inadequate Shock/Stimulation		
									A090801	Signal Artifact/Noise		
							F2203	Imaging Required	A070909	Oversensing		
									A071302	Inappropriate/Inadequate Shock/Stimulation		
									A090801	Signal Artifact/Noise		
17151	2025	All other reportable incidents	47263	Cardiac resynchronization therapy implantable	CT1685	Cardiac pacemakers	E1906	Unspecified Infection	A24	No Apparent Device Problem	24/01/2025	24/01/2025

				pacemaker			F08	Hospitalization or Prolonged Hospitalization	A24	No Apparent Device Problem	24/01/2025	24/01/2025
							F12	Serious Injury/ Illness/ Impairment	A24	No Apparent Device Problem	24/01/2025	24/01/2025
							F1903	Device Explantation	A24	No Apparent Device Problem	24/01/2025	24/01/2025
					CT257	Implantable cardiac pacemakers	E1906	Unspecified Infection	A24	No Apparent Device Problem	24/01/2025	24/01/2025
							F08	Hospitalization or Prolonged Hospitalization	A24	No Apparent Device Problem	24/01/2025	24/01/2025
							F12	Serious Injury/ Illness/ Impairment	A24	No Apparent Device Problem	24/01/2025	24/01/2025
							F1903	Device Explantation	A24	No Apparent Device Problem	24/01/2025	24/01/2025
					CT2866	Implantable cardiac electrical stimulators	E1906	Unspecified Infection	A24	No Apparent Device Problem	24/01/2025	24/01/2025
							F08	Hospitalization or Prolonged Hospitalization	A24	No Apparent Device Problem	24/01/2025	24/01/2025
							F12	Serious Injury/ Illness/ Impairment	A24	No Apparent Device Problem	24/01/2025	24/01/2025
							F1903	Device Explantation	A24	No Apparent Device Problem	24/01/2025	24/01/2025
17152	2025	All other reportable incidents	37265	Dual-chamber implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E210401	Shock from Patient Lead(s)	A0709	Device Sensing Problem		
									A071302	Inappropriate/Inadequate Shock/Stimulation		
							F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A0709	Device Sensing Problem		
									A071302	Inappropriate/Inadequate Shock/Stimulation		
							F1203	Life Threatening Illness or Injury	A0709	Device Sensing Problem		
									A071302	Inappropriate/Inadequate Shock/Stimulation		
							F2303	Medication Required	A0709	Device Sensing Problem		
									A071302	Inappropriate/Inadequate Shock/Stimulation		
							F2307	Reprogramming of Device	A0709	Device Sensing Problem		
									A071302	Inappropriate/Inadequate Shock/Stimulation		
17153	2025	All other reportable incidents	37265	Dual-chamber implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E210401	Shock from Patient Lead(s)	A071201	Failure to Convert Rhythm		
									A071302	Inappropriate/Inadequate Shock/Stimulation		
									A072201	High Impedance		
									A0908	Incorrect, Inadequate or Imprecise Result or Readings		
									A090801	Signal Artifact/Noise		
							F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A071201	Failure to Convert Rhythm		
									A071302	Inappropriate/Inadequate Shock/Stimulation		
									A072201	High Impedance		
									A0908	Incorrect, Inadequate or Imprecise Result or Readings		
									A090801	Signal Artifact/Noise		
							F15	Recognized Device or Procedural Complication	A071201	Failure to Convert Rhythm		
									A071302	Inappropriate/Inadequate Shock/Stimulation		
									A072201	High Impedance		
									A0908	Incorrect, Inadequate or Imprecise Result or Readings		
									A090801	Signal Artifact/Noise		
17154	2025	All other reportable incidents	47265	Dual-chamber implantable pacemaker, rate-responsive	CT1685	Cardiac pacemakers	E0601	Arrhythmia	A071102	Failure to Interrogate	17/06/2025	17/06/2025
									A071204	No Pacing	17/06/2025	17/06/2025
							E2311	Discomfort	A071102	Failure to Interrogate	17/06/2025	17/06/2025
									A071204	No Pacing	17/06/2025	17/06/2025
							F1901	Additional Surgery	A071102	Failure to Interrogate	17/06/2025	17/06/2025
									A071204	No Pacing	17/06/2025	17/06/2025
					CT257	Implantable cardiac pacemakers	E0601	Arrhythmia	A071102	Failure to Interrogate	17/06/2025	17/06/2025
									A071204	No Pacing	17/06/2025	17/06/2025
							E2311	Discomfort	A071102	Failure to Interrogate	17/06/2025	17/06/2025
									A071204	No Pacing	17/06/2025	17/06/2025
							F1901	Additional Surgery	A071102	Failure to Interrogate	17/06/2025	17/06/2025
									A071204	No Pacing	17/06/2025	17/06/2025
					CT2866	Implantable cardiac electrical stimulators	E0601	Arrhythmia	A071102	Failure to Interrogate	17/06/2025	17/06/2025
									A071204	No Pacing	17/06/2025	17/06/2025
							E2311	Discomfort	A071102	Failure to Interrogate	17/06/2025	17/06/2025
									A071204	No Pacing	17/06/2025	17/06/2025
							F1901	Additional Surgery	A071102	Failure to Interrogate	17/06/2025	17/06/2025
									A071204	No Pacing	17/06/2025	17/06/2025
17155	2025		47265	Dual-chamber implantable pacemaker, rate-responsive			E2403	No Clinical Signs, Symptoms or Conditions	A070909	Oversensing	19/06/2025	19/06/2025
									A071204	No Pacing	19/06/2025	19/06/2025
							F1901	Additional Surgery	A070909	Oversensing	19/06/2025	19/06/2025
									A071204	No Pacing	19/06/2025	19/06/2025
17156	2025	All other reportable incidents	35852	Single-chamber implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E060109	Tachycardia	A0709	Device Sensing Problem		
							E210401	Shock from Patient Lead(s)	A0709	Device Sensing Problem		
							F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A0709	Device Sensing Problem		
							F1203	Life Threatening Illness or Injury	A0709	Device Sensing Problem		
							F2307	Reprogramming of Device	A0709	Device Sensing Problem		
17157	2025		47265	Dual-chamber implantable pacemaker, rate-responsive	CT1685	Cardiac pacemakers	E2401	Insufficient Information	A02	Manufacturing, Packaging or Shipping Problem		
									A0203	Defective Device		
									A0705	Battery Problem		
							F19	Surgical Intervention	A02	Manufacturing, Packaging or Shipping Problem		
									A0203	Defective Device		
									A0705	Battery Problem		
							F1905	Device Revision or Replacement	A02	Manufacturing, Packaging or Shipping Problem		
									A0203	Defective Device		
									A0705	Battery Problem		
							F25	Unanticipated Adverse Device Effect	A02	Manufacturing, Packaging or Shipping Problem		
									A0203	Defective Device		
									A0705	Battery Problem		
					CT257	Implantable cardiac pacemakers	E2401	Insufficient Information	A02	Manufacturing, Packaging or Shipping Problem		
									A0203	Defective Device		
									A0705	Battery Problem		
							F19	Surgical Intervention	A02	Manufacturing, Packaging or Shipping Problem		
									A0203	Defective Device		
									A0705	Battery Problem		
							F1905	Device Revision or Replacement	A02	Manufacturing, Packaging or Shipping Problem		
									A0203	Defective Device		
									A0705	Battery Problem		
							F25	Unanticipated Adverse Device Effect	A02	Manufacturing, Packaging or Shipping Problem		

									A0203	Defective Device		
									A0705	Battery Problem		
					CT2866	Implantable cardiac electrical stimulators	E2401	Insufficient Information	A02	Manufacturing, Packaging or Shipping Problem		
									A0203	Defective Device		
									A0705	Battery Problem		
							F19	Surgical Intervention	A02	Manufacturing, Packaging or Shipping Problem		
									A0203	Defective Device		
									A0705	Battery Problem		
							F1905	Device Revision or Replacement	A02	Manufacturing, Packaging or Shipping Problem		
									A0203	Defective Device		
									A0705	Battery Problem		
							F25	Unanticipated Adverse Device Effect	A02	Manufacturing, Packaging or Shipping Problem		
									A0203	Defective Device		
									A0705	Battery Problem		
17158	2025	All other reportable incidents	35852	Single-chamber implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A0709	Device Sensing Problem		
							F26	No Health Consequences or Impact	A0709	Device Sensing Problem		
17159	2025	All other reportable incidents	37265	Dual-chamber implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A070909	Oversensing		
							F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A070909	Oversensing		
17160	2025	All other reportable incidents	37265	Dual-chamber implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E0601	Arrhythmia	A070910	Undersensing		
									A071202	Inaccurate Synchronization		
									A1301	Failure to Read Input Signal		
							F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A070910	Undersensing		
									A071202	Inaccurate Synchronization		
									A1301	Failure to Read Input Signal		
17161	2025	All other reportable incidents	37265	Dual-chamber implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A070101	Failure to Capture		
							F26	No Health Consequences or Impact	A070101	Failure to Capture		
17162	2025	All other reportable incidents	35852	Single-chamber implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A072001	Electromagnetic Interference		
							F12	Serious Injury/ Illness/ Impairment	A072001	Electromagnetic Interference		
							F2307	Reprogramming of Device	A072001	Electromagnetic Interference		
17163	2025	All other reportable incidents	47265	Dual-chamber implantable pacemaker, rate-responsive	CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A0722	Impedance Problem		
							F22	Unexpected Diagnostic Intervention	A0722	Impedance Problem		
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A0722	Impedance Problem		
							F22	Unexpected Diagnostic Intervention	A0722	Impedance Problem		
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A0722	Impedance Problem		
							F22	Unexpected Diagnostic Intervention	A0722	Impedance Problem		
17164	2025	All other reportable incidents	47205	Cardiac pulse generator programmer	CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A1305	Wireless Communication Problem		
							F26	No Health Consequences or Impact	A1305	Wireless Communication Problem		
17165	2025	All other reportable incidents	47267	Single-chamber implantable pacemaker, rate-responsive	CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A05	Mechanical Problem	20/06/2025	20/06/2025
							F26	No Health Consequences or Impact	A05	Mechanical Problem	20/06/2025	20/06/2025
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A05	Mechanical Problem	20/06/2025	20/06/2025
							F26	No Health Consequences or Impact	A05	Mechanical Problem	20/06/2025	20/06/2025
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A05	Mechanical Problem	20/06/2025	20/06/2025
							F26	No Health Consequences or Impact	A05	Mechanical Problem	20/06/2025	20/06/2025
17166	2025	All other reportable incidents	47263	Cardiac resynchronization therapy implantable pacemaker	CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A070102	High Capture Threshold		
									A072202	Low Impedance		
							F26	No Health Consequences or Impact	A070102	High Capture Threshold		
									A072202	Low Impedance		
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A070102	High Capture Threshold		
									A072202	Low Impedance		
							F26	No Health Consequences or Impact	A070102	High Capture Threshold		
									A072202	Low Impedance		
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A070102	High Capture Threshold		
									A072202	Low Impedance		
							F26	No Health Consequences or Impact	A070102	High Capture Threshold		
									A072202	Low Impedance		
17167	2025	All other reportable incidents	37265	Dual-chamber implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E2104	Electric Shock	A070910	Undersensing		
									A0712	Pacing Problem		
									A071302	Inappropriate/Inadequate Shock/Stimulation		
							F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A070910	Undersensing		
									A0712	Pacing Problem		
									A071302	Inappropriate/Inadequate Shock/Stimulation		
17168	2025	All other reportable incidents	47265	Dual-chamber implantable pacemaker, rate-responsive			E2103	Device Overstimulation of Tissue	A0908	Incorrect, Inadequate or Imprecise Result or Readings	04/07/2025	04/07/2025
							F08	Hospitalization or Prolonged Hospitalization	A0908	Incorrect, Inadequate or Imprecise Result or Readings	04/07/2025	04/07/2025
							F15	Recognized Device or Procedural Complication	A0908	Incorrect, Inadequate or Imprecise Result or Readings	04/07/2025	04/07/2025
							F1905	Device Revision or Replacement	A0908	Incorrect, Inadequate or Imprecise Result or Readings	04/07/2025	04/07/2025
17169	2025	All other reportable incidents	47265	Dual-chamber implantable pacemaker, rate-responsive	CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A070101	Failure to Capture	23/06/2025	23/06/2025
									A070909	Oversensing	23/06/2025	23/06/2025
							F1908	Prolonged Surgery	A070101	Failure to Capture	23/06/2025	23/06/2025
									A070909	Oversensing	23/06/2025	23/06/2025
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A070101	Failure to Capture	23/06/2025	23/06/2025
									A070909	Oversensing	23/06/2025	23/06/2025
							F1908	Prolonged Surgery	A070101	Failure to Capture	23/06/2025	23/06/2025
									A070909	Oversensing	23/06/2025	23/06/2025
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A070101	Failure to Capture	23/06/2025	23/06/2025
									A070909	Oversensing	23/06/2025	23/06/2025
							F1908	Prolonged Surgery	A070101	Failure to Capture	23/06/2025	23/06/2025
									A070909	Oversensing	23/06/2025	23/06/2025
17170	2025	All other reportable incidents	36106	Implantable pacemaker analyser, intraoperative	CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A1104	Computer Operating System Problem		
							F26	No Health Consequences or Impact	A1104	Computer Operating System Problem		
17171	2025	All other reportable incidents	47270	Cardiac resynchronization therapy implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A070909	Oversensing		
									A0712	Pacing Problem		
							F1001	Absence of Treatment	A070909	Oversensing		
									A0712	Pacing Problem		
17172	2025	All other reportable incidents	47265	Dual-chamber implantable pacemaker, rate-responsive			E2403	No Clinical Signs, Symptoms or Conditions	A0908	Incorrect, Inadequate or Imprecise Result or Readings	24/06/2025	24/06/2025

							F08	Hospitalization or Prolonged Hospitalization	A0908	Incorrect, Inadequate or Imprecise Result or Readings	24/06/2025	24/06/2025
							F1905	Device Revision or Replacement	A0908	Incorrect, Inadequate or Imprecise Result or Readings	24/06/2025	24/06/2025
17173	2025		17882	Physiologic-monitoring defibrillation system	CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A0203	Defective Device		
									A150105	Key or Button Unresponsive/not Working		
							F26	No Health Consequences or Impact	A0203	Defective Device		
									A150105	Key or Button Unresponsive/not Working		
17174	2025	All other reportable incidents	35852	Single-chamber implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A070909	Oversensing		
									A072001	Electromagnetic Interference		
							F26	No Health Consequences or Impact	A070909	Oversensing		
									A072001	Electromagnetic Interference		
17175	2025	All other reportable incidents	60789	Intracardiac pacemaker	CT1685	Cardiac pacemakers	E0503	Embolism/Embolus	A051201	Device Dislodged or Dislocated		
							F1908	Prolonged Surgery	A051201	Device Dislodged or Dislocated		
					CT257	Implantable cardiac pacemakers	E0503	Embolism/Embolus	A051201	Device Dislodged or Dislocated		
							F1908	Prolonged Surgery	A051201	Device Dislodged or Dislocated		
					CT2818	Catheterization/Cannulation/Fluid path devices	E0503	Embolism/Embolus	A051201	Device Dislodged or Dislocated		
							F1908	Prolonged Surgery	A051201	Device Dislodged or Dislocated		
					CT2866	Implantable cardiac electrical stimulators	E0503	Embolism/Embolus	A051201	Device Dislodged or Dislocated		
							F1908	Prolonged Surgery	A051201	Device Dislodged or Dislocated		
17176	2025	All other reportable incidents	47205	Cardiac pulse generator programmer	CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A0902	Display or Visual Feedback Problem		
							F26	No Health Consequences or Impact	A0902	Display or Visual Feedback Problem		
17177	2025		35852	Single-chamber implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E011903	Syncope/Fainting	A0706	Charging Problem		
									A071301	Failure to Deliver Shock/Stimulation		
							E0601	Arrhythmia	A0706	Charging Problem		
									A071301	Failure to Deliver Shock/Stimulation		
							F02	Death	A0706	Charging Problem		
									A071301	Failure to Deliver Shock/Stimulation		
17178	2025	All other reportable incidents	17882	Physiologic-monitoring defibrillation system	CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070803	Failure to Power Up		
							F27	Problem Identified During Non-Clinical Procedure	A070803	Failure to Power Up		
17179	2025	All other reportable incidents	35852	Single-chamber implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A010402	Migration		
							F2203	Imaging Required	A010402	Migration		
17180	2025	All other reportable incidents	47265	Dual-chamber implantable pacemaker, rate-responsive	CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A071102	Failure to Interrogate		
							F1901	Additional Surgery	A071102	Failure to Interrogate		
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A071102	Failure to Interrogate		
							F1901	Additional Surgery	A071102	Failure to Interrogate		
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A071102	Failure to Interrogate		
							F1901	Additional Surgery	A071102	Failure to Interrogate		
17181	2025	All other reportable incidents	35224	External pacemaker, epicardial pacing	CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A090206	No Display/Image		
							F26	No Health Consequences or Impact	A090206	No Display/Image		
17182	2025	All other reportable incidents	47265	Dual-chamber implantable pacemaker, rate-responsive	CT1685	Cardiac pacemakers	E060104	Bradycardia	A070103	Intermittent Capture		
							F08	Hospitalization or Prolonged Hospitalization	A070103	Intermittent Capture		
					CT257	Implantable cardiac pacemakers	E060104	Bradycardia	A070103	Intermittent Capture		
							F08	Hospitalization or Prolonged Hospitalization	A070103	Intermittent Capture		
					CT2866	Implantable cardiac electrical stimulators	E060104	Bradycardia	A070103	Intermittent Capture		
							F08	Hospitalization or Prolonged Hospitalization	A070103	Intermittent Capture		
17183	2025	All other reportable incidents	37265	Dual-chamber implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E0601	Arrhythmia	A070910	Undersensing		
							F23	Unexpected Medical Intervention	A070910	Undersensing		
17184	2025	All other reportable incidents	47270	Cardiac resynchronization therapy implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A070101	Failure to Capture		
									A070910	Undersensing		
									A0722	Impedance Problem		
							F26	No Health Consequences or Impact	A070101	Failure to Capture		
									A070910	Undersensing		
									A0722	Impedance Problem		
17185	2025	All other reportable incidents	17882	Physiologic-monitoring defibrillation system	CT1684	Cardiac pacemakers and associated devices	E0623	Unspecified Heart Problem	A0908	Incorrect, Inadequate or Imprecise Result or Readings		
							F05	Delay to Treatment/ Therapy	A0908	Incorrect, Inadequate or Imprecise Result or Readings		
17186	2025	All other reportable incidents	17882	Physiologic-monitoring defibrillation system	CT1684	Cardiac pacemakers and associated devices	E0623	Unspecified Heart Problem	A0908	Incorrect, Inadequate or Imprecise Result or Readings		
							F05	Delay to Treatment/ Therapy	A0908	Incorrect, Inadequate or Imprecise Result or Readings		
17187	2025	All other reportable incidents	35852	Single-chamber implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A090811	Unable to Obtain Readings		
							F26	No Health Consequences or Impact	A090811	Unable to Obtain Readings		
17188	2025	All other reportable incidents	63188	Left-ventricular wireless pacing system sensor/transmitter	CT145	Cardiovascular devices	E0623	Unspecified Heart Problem	A0701	Capturing Problem		
							F24	Insufficient Health Impact Information	A0701	Capturing Problem		
					CT1577	Wireless monitoring system transmitters	E0623	Unspecified Heart Problem	A0701	Capturing Problem		
							F24	Insufficient Health Impact Information	A0701	Capturing Problem		
					CT1684	Cardiac pacemakers and associated devices	E0623	Unspecified Heart Problem	A0701	Capturing Problem		
							F24	Insufficient Health Impact Information	A0701	Capturing Problem		
					CT2727	Active implant devices/ancillaries	E0623	Unspecified Heart Problem	A0701	Capturing Problem		
							F24	Insufficient Health Impact Information	A0701	Capturing Problem		
					CT2750	Implantable cardiac electrical stimulators and associated devices	E0623	Unspecified Heart Problem	A0701	Capturing Problem		
							F24	Insufficient Health Impact Information	A0701	Capturing Problem		
					CT2885	Physiological monitoring/measuring system transmitters/receivers/transceivers	E0623	Unspecified Heart Problem	A0701	Capturing Problem		
							F24	Insufficient Health Impact Information	A0701	Capturing Problem		
17189	2025	All other reportable incidents	60789	Intracardiac pacemaker	CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A070910	Undersensing		
							F12	Serious Injury/ Illness/ Impairment	A070910	Undersensing		
							F2307	Reprogramming of Device	A070910	Undersensing		
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A070910	Undersensing		
							F12	Serious Injury/ Illness/ Impairment	A070910	Undersensing		
							F2307	Reprogramming of Device	A070910	Undersensing		
					CT2818	Catheterization/Cannulation/Fluid path devices	E2403	No Clinical Signs, Symptoms or Conditions	A070910	Undersensing		
							F12	Serious Injury/ Illness/ Impairment	A070910	Undersensing		
							F2307	Reprogramming of Device	A070910	Undersensing		
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A070910	Undersensing		
							F12	Serious Injury/ Illness/ Impairment	A070910	Undersensing		
							F2307	Reprogramming of Device	A070910	Undersensing		
17190	2025	All other reportable incidents	60789	Intracardiac pacemaker	CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A070910	Undersensing		

							F12	Serious Injury/ Illness/ Impairment	A070910	Undersensing		
							F2307	Reprogramming of Device	A070910	Undersensing		
							F26	No Health Consequences or Impact	A070910	Undersensing		
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A070910	Undersensing		
							F12	Serious Injury/ Illness/ Impairment	A070910	Undersensing		
							F2307	Reprogramming of Device	A070910	Undersensing		
							F26	No Health Consequences or Impact	A070910	Undersensing		
					CT2818	Catheterization/Cannulation/Fluid path devices	E2403	No Clinical Signs, Symptoms or Conditions	A070910	Undersensing		
							F12	Serious Injury/ Illness/ Impairment	A070910	Undersensing		
							F2307	Reprogramming of Device	A070910	Undersensing		
							F26	No Health Consequences or Impact	A070910	Undersensing		
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A070910	Undersensing		
							F12	Serious Injury/ Illness/ Impairment	A070910	Undersensing		
							F2307	Reprogramming of Device	A070910	Undersensing		
							F26	No Health Consequences or Impact	A070910	Undersensing		
17191	2025	All other reportable incidents	47265	Dual-chamber implantable pacemaker, rate-responsive	CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
17192	2025	All other reportable incidents	47265	Dual-chamber implantable pacemaker, rate-responsive	CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A0712	Pacing Problem		
							F26	No Health Consequences or Impact	A0712	Pacing Problem		
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A0712	Pacing Problem		
							F26	No Health Consequences or Impact	A0712	Pacing Problem		
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A0712	Pacing Problem		
							F26	No Health Consequences or Impact	A0712	Pacing Problem		
17193	2025	All other reportable incidents	36106	Implantable pacemaker analyser, intraoperative	CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A09	Output Problem		
									A0902	Display or Visual Feedback Problem		
							F26	No Health Consequences or Impact	A09	Output Problem		
									A0902	Display or Visual Feedback Problem		
17194	2025	All other reportable incidents	61808	Vascular guide-catheter, reprocessed	CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A030202	Coagulation in Device or Device Ingredient		
							F26	No Health Consequences or Impact	A030202	Coagulation in Device or Device Ingredient		
17195	2025	All other reportable incidents	60789	Intracardiac pacemaker	CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A0504	Leak/Splash		
									A24	No Apparent Device Problem		
							F26	No Health Consequences or Impact	A0504	Leak/Splash		
									A24	No Apparent Device Problem		
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A0504	Leak/Splash		
									A24	No Apparent Device Problem		
							F26	No Health Consequences or Impact	A0504	Leak/Splash		
									A24	No Apparent Device Problem		
					CT2818	Catheterization/Cannulation/Fluid path devices	E2403	No Clinical Signs, Symptoms or Conditions	A0504	Leak/Splash		
									A24	No Apparent Device Problem		
							F26	No Health Consequences or Impact	A0504	Leak/Splash		
									A24	No Apparent Device Problem		
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A0504	Leak/Splash		
									A24	No Apparent Device Problem		
							F26	No Health Consequences or Impact	A0504	Leak/Splash		
									A24	No Apparent Device Problem		
17196	2025	All other reportable incidents	47263	Cardiac resynchronization therapy implantable pacemaker	CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A071205	Pacemaker Found in Back-Up Mode	09/07/2025	09/07/2025
									A0908	Incorrect, Inadequate or Imprecise Result or Readings	09/07/2025	09/07/2025
							F08	Hospitalization or Prolonged Hospitalization	A071205	Pacemaker Found in Back-Up Mode	09/07/2025	09/07/2025
									A0908	Incorrect, Inadequate or Imprecise Result or Readings	09/07/2025	09/07/2025
							F1905	Device Revision or Replacement	A071205	Pacemaker Found in Back-Up Mode	09/07/2025	09/07/2025
									A0908	Incorrect, Inadequate or Imprecise Result or Readings	09/07/2025	09/07/2025
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A071205	Pacemaker Found in Back-Up Mode	09/07/2025	09/07/2025
									A0908	Incorrect, Inadequate or Imprecise Result or Readings	09/07/2025	09/07/2025
							F08	Hospitalization or Prolonged Hospitalization	A071205	Pacemaker Found in Back-Up Mode	09/07/2025	09/07/2025
									A0908	Incorrect, Inadequate or Imprecise Result or Readings	09/07/2025	09/07/2025
							F1905	Device Revision or Replacement	A071205	Pacemaker Found in Back-Up Mode	09/07/2025	09/07/2025
									A0908	Incorrect, Inadequate or Imprecise Result or Readings	09/07/2025	09/07/2025
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A071205	Pacemaker Found in Back-Up Mode	09/07/2025	09/07/2025
									A0908	Incorrect, Inadequate or Imprecise Result or Readings	09/07/2025	09/07/2025
							F08	Hospitalization or Prolonged Hospitalization	A071205	Pacemaker Found in Back-Up Mode	09/07/2025	09/07/2025
									A0908	Incorrect, Inadequate or Imprecise Result or Readings	09/07/2025	09/07/2025
							F1905	Device Revision or Replacement	A071205	Pacemaker Found in Back-Up Mode	09/07/2025	09/07/2025
									A0908	Incorrect, Inadequate or Imprecise Result or Readings	09/07/2025	09/07/2025
17197	2025	All other reportable incidents	47265	Dual-chamber implantable pacemaker, rate-responsive	CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A070910	Undersensing		
							F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A070910	Undersensing		
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A070910	Undersensing		
							F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A070910	Undersensing		
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A070910	Undersensing		
							F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A070910	Undersensing		
17198	2025	All other reportable incidents	47267	Single-chamber implantable pacemaker, rate-responsive	CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A070909	Oversensing		
									A072201	High Impedance		

									A090801	Signal Artifact/Noise		
							F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A070909	Oversensing		
									A072201	High Impedance		
									A090801	Signal Artifact/Noise		
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A070909	Oversensing		
									A072201	High Impedance		
									A090801	Signal Artifact/Noise		
							F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A070909	Oversensing		
									A072201	High Impedance		
									A090801	Signal Artifact/Noise		
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A070909	Oversensing		
									A072201	High Impedance		
									A090801	Signal Artifact/Noise		
							F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A070909	Oversensing		
									A072201	High Impedance		
									A090801	Signal Artifact/Noise		
17199	2025	All other reportable incidents	47265	Dual-chamber implantable pacemaker, rate-responsive	CT1685	Cardiac pacemakers	E0119	Loss of Consciousness	A0712	Pacing Problem		
									A0908	Incorrect, Inadequate or Imprecise Result or Readings		
							E011903	Syncope/Fainting	A0712	Pacing Problem		
									A0908	Incorrect, Inadequate or Imprecise Result or Readings		
							F1001	Absence of Treatment	A0712	Pacing Problem		
									A0908	Incorrect, Inadequate or Imprecise Result or Readings		
					CT257	Implantable cardiac pacemakers	E0119	Loss of Consciousness	A0712	Pacing Problem		
									A0908	Incorrect, Inadequate or Imprecise Result or Readings		
							E011903	Syncope/Fainting	A0712	Pacing Problem		
									A0908	Incorrect, Inadequate or Imprecise Result or Readings		
							F1001	Absence of Treatment	A0712	Pacing Problem		
									A0908	Incorrect, Inadequate or Imprecise Result or Readings		
					CT2866	Implantable cardiac electrical stimulators	E0119	Loss of Consciousness	A0712	Pacing Problem		
									A0908	Incorrect, Inadequate or Imprecise Result or Readings		
							E011903	Syncope/Fainting	A0712	Pacing Problem		
									A0908	Incorrect, Inadequate or Imprecise Result or Readings		
							F1001	Absence of Treatment	A0712	Pacing Problem		
									A0908	Incorrect, Inadequate or Imprecise Result or Readings		
17200	2025	All other reportable incidents	47265	Dual-chamber implantable pacemaker, rate-responsive	CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A0712	Pacing Problem	11/04/2025	11/04/2025
									A072201	High Impedance	11/04/2025	11/04/2025
							F08	Hospitalization or Prolonged Hospitalization	A0712	Pacing Problem	11/04/2025	11/04/2025
									A072201	High Impedance	11/04/2025	11/04/2025
							F1905	Device Revision or Replacement	A0712	Pacing Problem	11/04/2025	11/04/2025
									A072201	High Impedance	11/04/2025	11/04/2025
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A0712	Pacing Problem	11/04/2025	11/04/2025
									A072201	High Impedance	11/04/2025	11/04/2025
							F08	Hospitalization or Prolonged Hospitalization	A0712	Pacing Problem	11/04/2025	11/04/2025
									A072201	High Impedance	11/04/2025	11/04/2025
							F1905	Device Revision or Replacement	A0712	Pacing Problem	11/04/2025	11/04/2025
									A072201	High Impedance	11/04/2025	11/04/2025
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A0712	Pacing Problem	11/04/2025	11/04/2025
									A072201	High Impedance	11/04/2025	11/04/2025
							F08	Hospitalization or Prolonged Hospitalization	A0712	Pacing Problem	11/04/2025	11/04/2025
									A072201	High Impedance	11/04/2025	11/04/2025
							F1905	Device Revision or Replacement	A0712	Pacing Problem	11/04/2025	11/04/2025
									A072201	High Impedance	11/04/2025	11/04/2025
17201	2025	All other reportable incidents	37265	Dual-chamber implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E210401	Shock from Patient Lead(s)	A071302	Inappropriate/Inadequate Shock/Stimulation		
							F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A071302	Inappropriate/Inadequate Shock/Stimulation		
17202	2025	All other reportable incidents	60956	Pacing/defibrillation lead tunneller	CT1684	Cardiac pacemakers and associated devices	E210401	Shock from Patient Lead(s)	A040101	Fracture	04/07/2025	04/07/2025
									A070909	Oversensing	04/07/2025	04/07/2025
									A071302	Inappropriate/Inadequate Shock/Stimulation	04/07/2025	04/07/2025
									A090801	Signal Artifact/Noise	04/07/2025	04/07/2025
							F05	Delay to Treatment/ Therapy	A040101	Fracture	04/07/2025	04/07/2025
									A070909	Oversensing	04/07/2025	04/07/2025
									A071302	Inappropriate/Inadequate Shock/Stimulation	04/07/2025	04/07/2025
									A090801	Signal Artifact/Noise	04/07/2025	04/07/2025
							F08	Hospitalization or Prolonged Hospitalization	A040101	Fracture	04/07/2025	04/07/2025
									A070909	Oversensing	04/07/2025	04/07/2025
									A071302	Inappropriate/Inadequate Shock/Stimulation	04/07/2025	04/07/2025
									A090801	Signal Artifact/Noise	04/07/2025	04/07/2025
							F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A040101	Fracture	04/07/2025	04/07/2025
									A070909	Oversensing	04/07/2025	04/07/2025
									A071302	Inappropriate/Inadequate Shock/Stimulation	04/07/2025	04/07/2025
									A090801	Signal Artifact/Noise	04/07/2025	04/07/2025
							F11	Non-Serious Injury/ Illness/ Impairment	A040101	Fracture	04/07/2025	04/07/2025
									A070909	Oversensing	04/07/2025	04/07/2025
									A071302	Inappropriate/Inadequate Shock/Stimulation	04/07/2025	04/07/2025
									A090801	Signal Artifact/Noise	04/07/2025	04/07/2025
							F1905	Device Revision or Replacement	A040101	Fracture	04/07/2025	04/07/2025
									A070909	Oversensing	04/07/2025	04/07/2025
									A071302	Inappropriate/Inadequate Shock/Stimulation	04/07/2025	04/07/2025
									A090801	Signal Artifact/Noise	04/07/2025	04/07/2025
							F23	Unexpected Medical Intervention	A040101	Fracture	04/07/2025	04/07/2025
									A070909	Oversensing	04/07/2025	04/07/2025
									A071302	Inappropriate/Inadequate Shock/Stimulation	04/07/2025	04/07/2025

									A090801	Signal Artifact/Noise	04/07/2025	04/07/2025
17203	2025	All other reportable incidents	35852	Single-chamber implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem	11/07/2025	11/07/2025
									A070504	Premature Discharge of Battery	11/07/2025	11/07/2025
							F08	Hospitalization or Prolonged Hospitalization	A0705	Battery Problem	11/07/2025	11/07/2025
									A070504	Premature Discharge of Battery	11/07/2025	11/07/2025
							F1905	Device Revision or Replacement	A0705	Battery Problem	11/07/2025	11/07/2025
									A070504	Premature Discharge of Battery	11/07/2025	11/07/2025
17204	2025	All other reportable incidents	36106	Implantable pacemaker analyser, intraoperative	CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A1104	Computer Operating System Problem		
							F23	Unexpected Medical Intervention	A1104	Computer Operating System Problem		
17205	2025	All other reportable incidents	36106	Implantable pacemaker analyser, intraoperative	CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A1104	Computer Operating System Problem		
							F23	Unexpected Medical Intervention	A1104	Computer Operating System Problem		
17206	2025	All other reportable incidents	36106	Implantable pacemaker analyser, intraoperative	CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A1104	Computer Operating System Problem		
							F23	Unexpected Medical Intervention	A1104	Computer Operating System Problem		
17207	2025	All other reportable incidents	36106	Implantable pacemaker analyser, intraoperative	CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A1104	Computer Operating System Problem		
							F23	Unexpected Medical Intervention	A1104	Computer Operating System Problem		
17208	2025	All other reportable incidents	47270	Cardiac resynchronization therapy implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E060109	Tachycardia	A0709	Device Sensing Problem		
									A0712	Pacing Problem		
							F05	Delay to Treatment/ Therapy	A0709	Device Sensing Problem		
									A0712	Pacing Problem		
							F1203	Life Threatening Illness or Injury	A0709	Device Sensing Problem		
									A0712	Pacing Problem		
17209	2025		47265	Dual-chamber implantable pacemaker, rate-responsive	CT1685	Cardiac pacemakers	E2311	Discomfort	A24	No Apparent Device Problem	03/07/2025	03/07/2025
							F1901	Additional Surgery	A24	No Apparent Device Problem	03/07/2025	03/07/2025
					CT257	Implantable cardiac pacemakers	E2311	Discomfort	A24	No Apparent Device Problem	03/07/2025	03/07/2025
							F1901	Additional Surgery	A24	No Apparent Device Problem	03/07/2025	03/07/2025
					CT2866	Implantable cardiac electrical stimulators	E2311	Discomfort	A24	No Apparent Device Problem	03/07/2025	03/07/2025
							F1901	Additional Surgery	A24	No Apparent Device Problem	03/07/2025	03/07/2025
17210	2025	All other reportable incidents	47205	Cardiac pulse generator programmer	CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A150105	Key or Button Unresponsive/not Working		
							F27	Problem Identified During Non-Clinical Procedure	A150105	Key or Button Unresponsive/not Working		
17211	2025	All other reportable incidents	17882	Physiologic-monitoring defibrillation system	CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A090811	Unable to Obtain Readings		
							F26	No Health Consequences or Impact	A090811	Unable to Obtain Readings		
17212	2025	All other reportable incidents	47267	Single-chamber implantable pacemaker, rate-responsive	CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A0203	Defective Device	21/04/2022	21/04/2022
									A070504	Premature Discharge of Battery	21/04/2022	21/04/2022
							F08	Hospitalization or Prolonged Hospitalization	A0203	Defective Device	21/04/2022	21/04/2022
									A070504	Premature Discharge of Battery	21/04/2022	21/04/2022
							F1905	Device Revision or Replacement	A0203	Defective Device	21/04/2022	21/04/2022
									A070504	Premature Discharge of Battery	21/04/2022	21/04/2022
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A0203	Defective Device	21/04/2022	21/04/2022
									A070504	Premature Discharge of Battery	21/04/2022	21/04/2022
							F08	Hospitalization or Prolonged Hospitalization	A0203	Defective Device	21/04/2022	21/04/2022
									A070504	Premature Discharge of Battery	21/04/2022	21/04/2022
							F1905	Device Revision or Replacement	A0203	Defective Device	21/04/2022	21/04/2022
									A070504	Premature Discharge of Battery	21/04/2022	21/04/2022
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A0203	Defective Device	21/04/2022	21/04/2022
									A070504	Premature Discharge of Battery	21/04/2022	21/04/2022
							F08	Hospitalization or Prolonged Hospitalization	A0203	Defective Device	21/04/2022	21/04/2022
									A070504	Premature Discharge of Battery	21/04/2022	21/04/2022
							F1905	Device Revision or Replacement	A0203	Defective Device	21/04/2022	21/04/2022
									A070504	Premature Discharge of Battery	21/04/2022	21/04/2022
17213	2025	All other reportable incidents	47270	Cardiac resynchronization therapy implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A13	Communication or Transmission Problem		
									A23	Use of Device Problem		
							F12	Serious Injury/ Illness/ Impairment	A13	Communication or Transmission Problem		
									A23	Use of Device Problem		
							F2307	Reprogramming of Device	A13	Communication or Transmission Problem		
									A23	Use of Device Problem		
17214	2025	All other reportable incidents	47205	Cardiac pulse generator programmer	CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A0508	Noise, Audible	31/07/2025	31/07/2025
									A070504	Premature Discharge of Battery	31/07/2025	31/07/2025
									A0908	Incorrect, Inadequate or Imprecise Result or Readings	31/07/2025	31/07/2025
							F08	Hospitalization or Prolonged Hospitalization	A0508	Noise, Audible	31/07/2025	31/07/2025
									A070504	Premature Discharge of Battery	31/07/2025	31/07/2025
									A0908	Incorrect, Inadequate or Imprecise Result or Readings	31/07/2025	31/07/2025
							F1905	Device Revision or Replacement	A0508	Noise, Audible	31/07/2025	31/07/2025
									A070504	Premature Discharge of Battery	31/07/2025	31/07/2025
									A0908	Incorrect, Inadequate or Imprecise Result or Readings	31/07/2025	31/07/2025
17215	2025	All other reportable incidents	35852	Single-chamber implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A0709	Device Sensing Problem		
							F12	Serious Injury/ Illness/ Impairment	A0709	Device Sensing Problem		
							F2307	Reprogramming of Device	A0709	Device Sensing Problem		
17216	2025		47270	Cardiac resynchronization therapy implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A25	No Apparent Adverse Event		
							F26	No Health Consequences or Impact	A25	No Apparent Adverse Event		
17217	2025	All other reportable incidents	60956	Pacing/defibrillation lead tunneller	CT1684	Cardiac pacemakers and associated devices	E210401	Shock from Patient Lead(s)	A070909	Oversensing		
									A071302	Inappropriate/Inadequate Shock/Stimulation		
									A090801	Signal Artifact/Noise		
							F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A070909	Oversensing		
									A071302	Inappropriate/Inadequate Shock/Stimulation		
									A090801	Signal Artifact/Noise		
							F2203	Imaging Required	A070909	Oversensing		
									A071302	Inappropriate/Inadequate Shock/Stimulation		
									A090801	Signal Artifact/Noise		
17218	2025	All other reportable incidents	47205	Cardiac pulse generator programmer	CT1684	Cardiac pacemakers and associated devices	E210401	Shock from Patient Lead(s)	A070909	Oversensing		
									A071302	Inappropriate/Inadequate Shock/Stimulation		
									A090801	Signal Artifact/Noise		
							F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A070909	Oversensing		
									A071302	Inappropriate/Inadequate Shock/Stimulation		
									A090801	Signal Artifact/Noise		
							F2203	Imaging Required	A070909	Oversensing		

									A071302	Inappropriate/Inadequate Shock/Stimulation		
									A090801	Signal Artifact/Noise		
17219	2025	All other reportable incidents	47265	Dual-chamber implantable pacemaker, rate-responsive			E2403	No Clinical Signs, Symptoms or Conditions	A071102	Failure to Interrogate	04/07/2025	04/07/2025
									A090801	Signal Artifact/Noise	04/07/2025	04/07/2025
							F1901	Additional Surgery	A071102	Failure to Interrogate	04/07/2025	04/07/2025
									A090801	Signal Artifact/Noise	04/07/2025	04/07/2025
17220	2025	All other reportable incidents	36106	Implantable pacemaker analyser, intraoperative	CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A09	Output Problem		
									A1107	Data Problem		
							F26	No Health Consequences or Impact	A09	Output Problem		
									A1107	Data Problem		
17221	2025	All other reportable incidents	36106	Implantable pacemaker analyser, intraoperative	CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A09	Output Problem		
							F26	No Health Consequences or Impact	A09	Output Problem		
17222	2025	Unanticipated serious deterioration in state of health	60789	Intracardiac pacemaker	CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A0701	Capturing Problem		
									A070102	High Capture Threshold		
									A070103	Intermittent Capture		
							F12	Serious Injury/ Illness/ Impairment	A0701	Capturing Problem		
									A070102	High Capture Threshold		
									A070103	Intermittent Capture		
							F2307	Reprogramming of Device	A0701	Capturing Problem		
									A070102	High Capture Threshold		
									A070103	Intermittent Capture		
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A0701	Capturing Problem		
									A070102	High Capture Threshold		
									A070103	Intermittent Capture		
							F12	Serious Injury/ Illness/ Impairment	A0701	Capturing Problem		
									A070102	High Capture Threshold		
									A070103	Intermittent Capture		
							F2307	Reprogramming of Device	A0701	Capturing Problem		
									A070102	High Capture Threshold		
									A070103	Intermittent Capture		
					CT2818	Catheterization/Cannulation/Fluid path devices	E2403	No Clinical Signs, Symptoms or Conditions	A0701	Capturing Problem		
									A070102	High Capture Threshold		
									A070103	Intermittent Capture		
							F12	Serious Injury/ Illness/ Impairment	A0701	Capturing Problem		
									A070102	High Capture Threshold		
									A070103	Intermittent Capture		
							F2307	Reprogramming of Device	A0701	Capturing Problem		
									A070102	High Capture Threshold		
									A070103	Intermittent Capture		
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A0701	Capturing Problem		
									A070102	High Capture Threshold		
									A070103	Intermittent Capture		
							F12	Serious Injury/ Illness/ Impairment	A0701	Capturing Problem		
									A070102	High Capture Threshold		
									A070103	Intermittent Capture		
							F2307	Reprogramming of Device	A0701	Capturing Problem		
									A070102	High Capture Threshold		
									A070103	Intermittent Capture		
17223	2025	All other reportable incidents	47267	Single-chamber implantable pacemaker, rate-responsive	CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
17224	2025	All other reportable incidents	37265	Dual-chamber implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem	05/08/2025	05/08/2025
									A070504	Premature Discharge of Battery	05/08/2025	05/08/2025
									A0908	Incorrect, Inadequate or Imprecise Result or Readings	05/08/2025	05/08/2025
							F08	Hospitalization or Prolonged Hospitalization	A0705	Battery Problem	05/08/2025	05/08/2025
									A070504	Premature Discharge of Battery	05/08/2025	05/08/2025
									A0908	Incorrect, Inadequate or Imprecise Result or Readings	05/08/2025	05/08/2025
							F1905	Device Revision or Replacement	A0705	Battery Problem	05/08/2025	05/08/2025
									A070504	Premature Discharge of Battery	05/08/2025	05/08/2025
									A0908	Incorrect, Inadequate or Imprecise Result or Readings	05/08/2025	05/08/2025
17225	2025	All other reportable incidents	17882	Physiologic-monitoring defibrillation system	CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A11	Computer Software Problem		
							F27	Problem Identified During Non-Clinical Procedure	A11	Computer Software Problem		
17226	2025	All other reportable incidents	47263	Cardiac resynchronization therapy implantable pacemaker	CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A070102	High Capture Threshold		
									A070504	Premature Discharge of Battery		
									A072201	High Impedance		
									A1301	Failure to Read Input Signal		
							F26	No Health Consequences or Impact	A070102	High Capture Threshold		
									A070504	Premature Discharge of Battery		
									A072201	High Impedance		
									A1301	Failure to Read Input Signal		
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A070102	High Capture Threshold		
									A070504	Premature Discharge of Battery		
									A072201	High Impedance		
									A1301	Failure to Read Input Signal		
							F26	No Health Consequences or Impact	A070102	High Capture Threshold		
									A070504	Premature Discharge of Battery		
									A072201	High Impedance		
									A1301	Failure to Read Input Signal		
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A070102	High Capture Threshold		
									A070504	Premature Discharge of Battery		
									A072201	High Impedance		
									A1301	Failure to Read Input Signal		

							F26	No Health Consequences or Impact	A070102	High Capture Threshold		
									A070504	Premature Discharge of Battery		
									A072201	High Impedance		
									A1301	Failure to Read Input Signal		
17227	2025	Unanticipated serious deterioration in state of health	47267	Single-chamber implantable pacemaker, rate-responsive	CT1685	Cardiac pacemakers	E0610	Endocarditis	A24	No Apparent Device Problem		
							F08	Hospitalization or Prolonged Hospitalization	A24	No Apparent Device Problem		
							F12	Serious Injury/ Illness/ Impairment	A24	No Apparent Device Problem		
							F1905	Device Revision or Replacement	A24	No Apparent Device Problem		
							F2303	Medication Required	A24	No Apparent Device Problem		
					CT257	Implantable cardiac pacemakers	E0610	Endocarditis	A24	No Apparent Device Problem		
							F08	Hospitalization or Prolonged Hospitalization	A24	No Apparent Device Problem		
							F12	Serious Injury/ Illness/ Impairment	A24	No Apparent Device Problem		
							F1905	Device Revision or Replacement	A24	No Apparent Device Problem		
							F2303	Medication Required	A24	No Apparent Device Problem		
					CT2866	Implantable cardiac electrical stimulators	E0610	Endocarditis	A24	No Apparent Device Problem		
							F08	Hospitalization or Prolonged Hospitalization	A24	No Apparent Device Problem		
							F12	Serious Injury/ Illness/ Impairment	A24	No Apparent Device Problem		
							F1905	Device Revision or Replacement	A24	No Apparent Device Problem		
							F2303	Medication Required	A24	No Apparent Device Problem		
17228	2025	All other reportable incidents	47265	Dual-chamber implantable pacemaker, rate-responsive	CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A05	Mechanical Problem	10/07/2025	10/07/2025
							F26	No Health Consequences or Impact	A05	Mechanical Problem	10/07/2025	10/07/2025
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A05	Mechanical Problem	10/07/2025	10/07/2025
							F26	No Health Consequences or Impact	A05	Mechanical Problem	10/07/2025	10/07/2025
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A05	Mechanical Problem	10/07/2025	10/07/2025
							F26	No Health Consequences or Impact	A05	Mechanical Problem	10/07/2025	10/07/2025
17229	2025		47270	Cardiac resynchronization therapy implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A070101	Failure to Capture	18/08/2025	18/08/2025
									A072201	High Impedance	18/08/2025	18/08/2025
							F1901	Additional Surgery	A070101	Failure to Capture	18/08/2025	18/08/2025
									A072201	High Impedance	18/08/2025	18/08/2025
17230	2025	All other reportable incidents	35852	Single-chamber implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E210401	Shock from Patient Lead(s)	A0709	Device Sensing Problem		
									A071302	Inappropriate/Inadequate Shock/Stimulation		
							F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A0709	Device Sensing Problem		
									A071302	Inappropriate/Inadequate Shock/Stimulation		
							F1203	Life Threatening Illness or Injury	A0709	Device Sensing Problem		
									A071302	Inappropriate/Inadequate Shock/Stimulation		
							F2307	Reprogramming of Device	A0709	Device Sensing Problem		
									A071302	Inappropriate/Inadequate Shock/Stimulation		
17231	2025	All other reportable incidents	37265	Dual-chamber implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A0712	Pacing Problem		
							F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A0712	Pacing Problem		
17232	2025	All other reportable incidents	47270	Cardiac resynchronization therapy implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E210401	Shock from Patient Lead(s)	A0712	Pacing Problem		
									A071302	Inappropriate/Inadequate Shock/Stimulation		
							F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A0712	Pacing Problem		
									A071302	Inappropriate/Inadequate Shock/Stimulation		
17233	2025	All other reportable incidents	36106	Implantable pacemaker analyser, intraoperative	CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A0709	Device Sensing Problem		
									A1104	Computer Operating System Problem		
							F26	No Health Consequences or Impact	A0709	Device Sensing Problem		
									A1104	Computer Operating System Problem		
17234	2025	All other reportable incidents	36106	Implantable pacemaker analyser, intraoperative	CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A0709	Device Sensing Problem		
									A1104	Computer Operating System Problem		
							F26	No Health Consequences or Impact	A0709	Device Sensing Problem		
									A1104	Computer Operating System Problem		
17235	2025	Unanticipated serious deterioration in state of health	47265	Dual-chamber implantable pacemaker, rate-responsive	CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A0712	Pacing Problem		
							F12	Serious Injury/ Illness/ Impairment	A0712	Pacing Problem		
							F2307	Reprogramming of Device	A0712	Pacing Problem		
							F26	No Health Consequences or Impact	A0712	Pacing Problem		
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A0712	Pacing Problem		
							F12	Serious Injury/ Illness/ Impairment	A0712	Pacing Problem		
							F2307	Reprogramming of Device	A0712	Pacing Problem		
							F26	No Health Consequences or Impact	A0712	Pacing Problem		
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A0712	Pacing Problem		
							F12	Serious Injury/ Illness/ Impairment	A0712	Pacing Problem		
							F2307	Reprogramming of Device	A0712	Pacing Problem		
							F26	No Health Consequences or Impact	A0712	Pacing Problem		
17236	2025	All other reportable incidents	37265	Dual-chamber implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A070101	Failure to Capture		
									A070909	Oversensing		
							F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A070101	Failure to Capture		
									A070909	Oversensing		
17237	2025	All other reportable incidents	37265	Dual-chamber implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A0712	Pacing Problem	03/07/2025	03/07/2025
							F08	Hospitalization or Prolonged Hospitalization	A0712	Pacing Problem	03/07/2025	03/07/2025
							F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A0712	Pacing Problem	03/07/2025	03/07/2025
							F1905	Device Revision or Replacement	A0712	Pacing Problem	03/07/2025	03/07/2025
17238	2025	All other reportable incidents	35852	Single-chamber implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem	28/08/2025	28/08/2025
									A070504	Premature Discharge of Battery	28/08/2025	28/08/2025
									A0908	Incorrect, Inadequate or Imprecise Result or Readings	28/08/2025	28/08/2025
							F08	Hospitalization or Prolonged Hospitalization	A0705	Battery Problem	28/08/2025	28/08/2025
									A070504	Premature Discharge of Battery	28/08/2025	28/08/2025
									A0908	Incorrect, Inadequate or Imprecise Result or Readings	28/08/2025	28/08/2025
							F1905	Device Revision or Replacement	A0705	Battery Problem	28/08/2025	28/08/2025
									A070504	Premature Discharge of Battery	28/08/2025	28/08/2025
									A0908	Incorrect, Inadequate or Imprecise Result or Readings	28/08/2025	28/08/2025
17239	2025	All other reportable incidents	37265	Dual-chamber implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	07/07/2025	07/07/2025
							F1901	Additional Surgery	A070504	Premature Discharge of Battery	07/07/2025	07/07/2025
17240	2025	All other reportable incidents	37265	Dual-chamber implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A0712	Pacing Problem		
							F26	No Health Consequences or Impact	A0712	Pacing Problem		
17241	2025	All other reportable incidents	47270	Cardiac resynchronization therapy implantable	CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A070909	Oversensing		

				defibrillator			F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A070909	Oversensing			
17242	2025	All other reportable incidents	46717	Pacing/defibrillation lead stylet	CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A150208	Entrapment of Device			
							F26	No Health Consequences or Impact	A150208	Entrapment of Device			
17243	2025	All other reportable incidents	47270	Cardiac resynchronization therapy implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A1205	Loose or Intermittent Connection	17/07/2025	17/07/2025	
							F1901	Additional Surgery	A1205	Loose or Intermittent Connection	17/07/2025	17/07/2025	
17244	2025		18145	Cardiac myoplasty electrical stimulation system	CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A040101	Fracture			
									A050401	Fluid/Blood Leak			
							F1905	Device Revision or Replacement	A040101	Fracture			
									A050401	Fluid/Blood Leak			
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A040101	Fracture			
									A050401	Fluid/Blood Leak			
							F1905	Device Revision or Replacement	A040101	Fracture			
									A050401	Fluid/Blood Leak			
17245	2025	All other reportable incidents	47270	Cardiac resynchronization therapy implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E2401	Insufficient Information	A071301	Failure to Deliver Shock/Stimulation			
							F24	Insufficient Health Impact Information	A071301	Failure to Deliver Shock/Stimulation			
17246	2025	Death	37265	Dual-chamber implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E020201	Anxiety	A071302	Inappropriate/Inadequate Shock/Stimulation			
							E210401	Shock from Patient Lead(s)	A071302	Inappropriate/Inadequate Shock/Stimulation			
							E233001	Chest Pain	A071302	Inappropriate/Inadequate Shock/Stimulation			
							F02	Death	A071302	Inappropriate/Inadequate Shock/Stimulation			
							F08	Hospitalization or Prolonged Hospitalization	A071302	Inappropriate/Inadequate Shock/Stimulation			
							F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A071302	Inappropriate/Inadequate Shock/Stimulation			
17247	2025	All other reportable incidents	47270	Cardiac resynchronization therapy implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A0712	Pacing Problem			
							F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A0712	Pacing Problem			
17248	2025	All other reportable incidents	47270	Cardiac resynchronization therapy implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A070101	Failure to Capture			
									A070910	Undersensing			
									A0712	Pacing Problem			
							F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A070101	Failure to Capture			
									A070910	Undersensing			
									A0712	Pacing Problem			
							F1001	Absence of Treatment	A070101	Failure to Capture			
									A070910	Undersensing			
									A0712	Pacing Problem			
							F15	Recognized Device or Procedural Complication	A070101	Failure to Capture			
									A070910	Undersensing			
									A0712	Pacing Problem			
17249	2025	All other reportable incidents	47265	Dual-chamber implantable pacemaker, rate-responsive			E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	04/07/2025	04/07/2025	
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	04/07/2025	04/07/2025	
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	04/07/2025	04/07/2025	
17250	2025	All other reportable incidents	47263	Cardiac resynchronization therapy implantable pacemaker	CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A070909	Oversensing			
									A0712	Pacing Problem			
									A1107	Data Problem			
							F1001	Absence of Treatment	A070909	Oversensing			
									A0712	Pacing Problem			
									A1107	Data Problem			
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A070909	Oversensing			
									A0712	Pacing Problem			
									A1107	Data Problem			
							F1001	Absence of Treatment	A070909	Oversensing			
									A0712	Pacing Problem			
									A1107	Data Problem			
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A070909	Oversensing			
									A0712	Pacing Problem			
									A1107	Data Problem			
							F1001	Absence of Treatment	A070909	Oversensing			
									A0712	Pacing Problem			
									A1107	Data Problem			
17251	2025		46717	Pacing/defibrillation lead stylet	CT1684	Cardiac pacemakers and associated devices	E0119	Loss of Consciousness	A0709	Device Sensing Problem			
									A070909	Oversensing			
									A071204	No Pacing			
									A150207	Difficult to Remove			
							E011903	Syncope/Fainting	A0709	Device Sensing Problem			
									A070909	Oversensing			
									A071204	No Pacing			
									A150207	Difficult to Remove			
							E020203	Distress	A0709	Device Sensing Problem			
									A070909	Oversensing			
									A071204	No Pacing			
									A150207	Difficult to Remove			
							F02	Death	A0709	Device Sensing Problem			
									A070909	Oversensing			
									A071204	No Pacing			
									A150207	Difficult to Remove			
							F08	Hospitalization or Prolonged Hospitalization	A0709	Device Sensing Problem			
									A070909	Oversensing			
									A071204	No Pacing			
									A150207	Difficult to Remove			
							F1903	Device Explantation	A0709	Device Sensing Problem			
									A070909	Oversensing			
									A071204	No Pacing			
									A150207	Difficult to Remove			
							F21	Unexpected Deterioration	A0709	Device Sensing Problem			
									A070909	Oversensing			
									A071204	No Pacing			
									A150207	Difficult to Remove			
17252	2025	Unanticipated serious deterioration in state of health	47265	Dual-chamber implantable pacemaker, rate-responsive	CT1685	Cardiac pacemakers	E0112	Dizziness	A071209	Extracardiac Stimulation			
									A072001	Electromagnetic Interference			
							E011903	Syncope/Fainting	A071209	Extracardiac Stimulation			
									A072001	Electromagnetic Interference			

							E0601	Arrhythmia	A071209	Extracardiac Stimulation		
									A072001	Electromagnetic Interference		
							E2103	Device Overstimulation of Tissue	A071209	Extracardiac Stimulation		
									A072001	Electromagnetic Interference		
							F1203	Life Threatening Illness or Injury	A071209	Extracardiac Stimulation		
									A072001	Electromagnetic Interference		
					CT257	Implantable cardiac pacemakers	E0112	Dizziness	A071209	Extracardiac Stimulation		
									A072001	Electromagnetic Interference		
							E011903	Syncope/Fainting	A071209	Extracardiac Stimulation		
									A072001	Electromagnetic Interference		
							E0601	Arrhythmia	A071209	Extracardiac Stimulation		
									A072001	Electromagnetic Interference		
							E2103	Device Overstimulation of Tissue	A071209	Extracardiac Stimulation		
									A072001	Electromagnetic Interference		
							F1203	Life Threatening Illness or Injury	A071209	Extracardiac Stimulation		
									A072001	Electromagnetic Interference		
					CT2866	Implantable cardiac electrical stimulators	E0112	Dizziness	A071209	Extracardiac Stimulation		
									A072001	Electromagnetic Interference		
							E011903	Syncope/Fainting	A071209	Extracardiac Stimulation		
									A072001	Electromagnetic Interference		
							E0601	Arrhythmia	A071209	Extracardiac Stimulation		
									A072001	Electromagnetic Interference		
							E2103	Device Overstimulation of Tissue	A071209	Extracardiac Stimulation		
									A072001	Electromagnetic Interference		
							F1203	Life Threatening Illness or Injury	A071209	Extracardiac Stimulation		
									A072001	Electromagnetic Interference		
17253	2025	All other reportable incidents	47267	Single-chamber implantable pacemaker, rate-responsive	CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
17254	2025	All other reportable incidents	17882	Physiologic-monitoring defibrillation system	CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A07	Electrical /Electronic Property Problem		
							F27	Problem Identified During Non-Clinical Procedure	A07	Electrical /Electronic Property Problem		
17255	2025	All other reportable incidents	47270	Cardiac resynchronization therapy implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A070909	Oversensing		
									A0712	Pacing Problem		
									A072201	High Impedance		
									A1301	Failure to Read Input Signal		
							F08	Hospitalization or Prolonged Hospitalization	A070909	Oversensing		
									A0712	Pacing Problem		
									A072201	High Impedance		
									A1301	Failure to Read Input Signal		
							F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A070909	Oversensing		
									A0712	Pacing Problem		
									A072201	High Impedance		
									A1301	Failure to Read Input Signal		
							F23	Unexpected Medical Intervention	A070909	Oversensing		
									A0712	Pacing Problem		
									A072201	High Impedance		
									A1301	Failure to Read Input Signal		
17256	2025	All other reportable incidents	47270	Cardiac resynchronization therapy implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A070910	Undersensing		
							F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A070910	Undersensing		
17257	2025	All other reportable incidents	47265	Dual-chamber implantable pacemaker, rate-responsive	CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A0908	Incorrect, Inadequate or Imprecise Result or Readings		
							F26	No Health Consequences or Impact	A0908	Incorrect, Inadequate or Imprecise Result or Readings		
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A0908	Incorrect, Inadequate or Imprecise Result or Readings		
							F26	No Health Consequences or Impact	A0908	Incorrect, Inadequate or Imprecise Result or Readings		
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A0908	Incorrect, Inadequate or Imprecise Result or Readings		
							F26	No Health Consequences or Impact	A0908	Incorrect, Inadequate or Imprecise Result or Readings		
17258	2025	All other reportable incidents	47265	Dual-chamber implantable pacemaker, rate-responsive	CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A071102	Failure to Interrogate	11/07/2025	11/07/2025
							F1901	Additional Surgery	A071102	Failure to Interrogate	11/07/2025	11/07/2025
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A071102	Failure to Interrogate	11/07/2025	11/07/2025
							F1901	Additional Surgery	A071102	Failure to Interrogate	11/07/2025	11/07/2025
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A071102	Failure to Interrogate	11/07/2025	11/07/2025
							F1901	Additional Surgery	A071102	Failure to Interrogate	11/07/2025	11/07/2025
17259	2025	All other reportable incidents	47265	Dual-chamber implantable pacemaker, rate-responsive			E0112	Dizziness	A0908	Incorrect, Inadequate or Imprecise Result or Readings	15/07/2025	15/07/2025
							E0623	Unspecified Heart Problem	A0908	Incorrect, Inadequate or Imprecise Result or Readings	15/07/2025	15/07/2025
							E0717	Dyspnea	A0908	Incorrect, Inadequate or Imprecise Result or Readings	15/07/2025	15/07/2025
							F08	Hospitalization or Prolonged Hospitalization	A0908	Incorrect, Inadequate or Imprecise Result or Readings	15/07/2025	15/07/2025
							F1905	Device Revision or Replacement	A0908	Incorrect, Inadequate or Imprecise Result or Readings	15/07/2025	15/07/2025
17260	2025	All other reportable incidents	60789	Intracardiac pacemaker	CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A070102	High Capture Threshold		
									A090805	Incorrect Measurement		
							F12	Serious Injury/ Illness/ Impairment	A070102	High Capture Threshold		
									A090805	Incorrect Measurement		

							F2307	Reprogramming of Device	A070102	High Capture Threshold		
									A090805	Incorrect Measurement		
							F26	No Health Consequences or Impact	A070102	High Capture Threshold		
									A090805	Incorrect Measurement		
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A070102	High Capture Threshold		
									A090805	Incorrect Measurement		
							F12	Serious Injury/ Illness/ Impairment	A070102	High Capture Threshold		
									A090805	Incorrect Measurement		
							F2307	Reprogramming of Device	A070102	High Capture Threshold		
									A090805	Incorrect Measurement		
							F26	No Health Consequences or Impact	A070102	High Capture Threshold		
									A090805	Incorrect Measurement		
					CT2818	Catheterization/Cannulation/Fluid path devices	E2403	No Clinical Signs, Symptoms or Conditions	A070102	High Capture Threshold		
									A090805	Incorrect Measurement		
							F12	Serious Injury/ Illness/ Impairment	A070102	High Capture Threshold		
									A090805	Incorrect Measurement		
							F2307	Reprogramming of Device	A070102	High Capture Threshold		
									A090805	Incorrect Measurement		
							F26	No Health Consequences or Impact	A070102	High Capture Threshold		
									A090805	Incorrect Measurement		
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A070102	High Capture Threshold		
									A090805	Incorrect Measurement		
							F12	Serious Injury/ Illness/ Impairment	A070102	High Capture Threshold		
									A090805	Incorrect Measurement		
							F2307	Reprogramming of Device	A070102	High Capture Threshold		
									A090805	Incorrect Measurement		
							F26	No Health Consequences or Impact	A070102	High Capture Threshold		
									A090805	Incorrect Measurement		
17261	2025		47270	Cardiac resynchronization therapy implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem		
							F26	No Health Consequences or Impact	A0705	Battery Problem		
17262	2025	All other reportable incidents	47265	Dual-chamber implantable pacemaker, rate-responsive	CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A0712	Pacing Problem	18/07/2025	18/07/2025
									A1605	Reset Problem	18/07/2025	18/07/2025
							F08	Hospitalization or Prolonged Hospitalization	A0712	Pacing Problem	18/07/2025	18/07/2025
									A1605	Reset Problem	18/07/2025	18/07/2025
							F12	Serious Injury/ Illness/ Impairment	A0712	Pacing Problem	18/07/2025	18/07/2025
									A1605	Reset Problem	18/07/2025	18/07/2025
							F1905	Device Revision or Replacement	A0712	Pacing Problem	18/07/2025	18/07/2025
									A1605	Reset Problem	18/07/2025	18/07/2025
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A0712	Pacing Problem	18/07/2025	18/07/2025
									A1605	Reset Problem	18/07/2025	18/07/2025
							F08	Hospitalization or Prolonged Hospitalization	A0712	Pacing Problem	18/07/2025	18/07/2025
									A1605	Reset Problem	18/07/2025	18/07/2025
							F12	Serious Injury/ Illness/ Impairment	A0712	Pacing Problem	18/07/2025	18/07/2025
									A1605	Reset Problem	18/07/2025	18/07/2025
							F1905	Device Revision or Replacement	A0712	Pacing Problem	18/07/2025	18/07/2025
									A1605	Reset Problem	18/07/2025	18/07/2025
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A0712	Pacing Problem	18/07/2025	18/07/2025
									A1605	Reset Problem	18/07/2025	18/07/2025
							F08	Hospitalization or Prolonged Hospitalization	A0712	Pacing Problem	18/07/2025	18/07/2025
									A1605	Reset Problem	18/07/2025	18/07/2025
							F12	Serious Injury/ Illness/ Impairment	A0712	Pacing Problem	18/07/2025	18/07/2025
									A1605	Reset Problem	18/07/2025	18/07/2025
							F1905	Device Revision or Replacement	A0712	Pacing Problem	18/07/2025	18/07/2025
									A1605	Reset Problem	18/07/2025	18/07/2025
17263	2025	All other reportable incidents	47270	Cardiac resynchronization therapy implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A070909	Oversensing		
							F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A070909	Oversensing		
17264	2025	All other reportable incidents	47265	Dual-chamber implantable pacemaker, rate-responsive			E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	07/07/2025	07/07/2025
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	07/07/2025	07/07/2025
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	07/07/2025	07/07/2025
17265	2025	All other reportable incidents	47270	Cardiac resynchronization therapy implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A0709	Device Sensing Problem		
									A0712	Pacing Problem		
							F26	No Health Consequences or Impact	A0709	Device Sensing Problem		
									A0712	Pacing Problem		
17266	2025	All other reportable incidents	17882	Physiologic-monitoring defibrillation system	CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A090811	Unable to Obtain Readings		
							F27	Problem Identified During Non-Clinical Procedure	A090811	Unable to Obtain Readings		
17267	2025	All other reportable incidents	35852	Single-chamber implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A072201	High Impedance		
							F26	No Health Consequences or Impact	A072201	High Impedance		
17268	2025	All other reportable incidents	47270	Cardiac resynchronization therapy implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E0119	Loss of Consciousness	A070909	Oversensing		
									A071201	Failure to Convert Rhythm		
									A0713	Defibrillation/Stimulation Problem		
									A072202	Low Impedance		
									A090801	Signal Artifact/Noise		
									A0909	Unexpected Therapeutic Results		
							E011903	Syncope/Fainting	A070909	Oversensing		
									A071201	Failure to Convert Rhythm		
									A0713	Defibrillation/Stimulation Problem		
									A072202	Low Impedance		
									A090801	Signal Artifact/Noise		
									A0909	Unexpected Therapeutic Results		
							E2104	Electric Shock	A070909	Oversensing		
									A071201	Failure to Convert Rhythm		
									A0713	Defibrillation/Stimulation Problem		
									A072202	Low Impedance		
									A090801	Signal Artifact/Noise		
									A0909	Unexpected Therapeutic Results		
							F05	Delay to Treatment/ Therapy	A070909	Oversensing		
									A071201	Failure to Convert Rhythm		
									A0713	Defibrillation/Stimulation Problem		

									A072202	Low Impedance			
									A090801	Signal Artifact/Noise			
									A0909	Unexpected Therapeutic Results			
							F08	Hospitalization or Prolonged Hospitalization	A070909	Oversensing			
									A071201	Failure to Convert Rhythm			
									A0713	Defibrillation/Stimulation Problem			
									A072202	Low Impedance			
									A090801	Signal Artifact/Noise			
									A0909	Unexpected Therapeutic Results			
							F1905	Device Revision or Replacement	A070909	Oversensing			
									A071201	Failure to Convert Rhythm			
									A0713	Defibrillation/Stimulation Problem			
									A072202	Low Impedance			
									A090801	Signal Artifact/Noise			
									A0909	Unexpected Therapeutic Results			
17269	2025	All other reportable incidents	35852	Single-chamber implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A160601	Premature Elective Replacement Indicator	06/11/2025	06/11/2025
							F1905	Device Revision or Replacement	A160601	Premature Elective Replacement Indicator	06/11/2025	06/11/2025
17270	2025	All other reportable incidents	47265	Dual-chamber implantable pacemaker, rate-responsive	CT1685	Cardiac pacemakers	E0601	Arrhythmia	A24	No Apparent Device Problem	15/07/2025	15/07/2025
							F1901	Additional Surgery	A24	No Apparent Device Problem	15/07/2025	15/07/2025
					CT257	Implantable cardiac pacemakers	E0601	Arrhythmia	A24	No Apparent Device Problem	15/07/2025	15/07/2025
							F1901	Additional Surgery	A24	No Apparent Device Problem	15/07/2025	15/07/2025
					CT2866	Implantable cardiac electrical stimulators	E0601	Arrhythmia	A24	No Apparent Device Problem	15/07/2025	15/07/2025
							F1901	Additional Surgery	A24	No Apparent Device Problem	15/07/2025	15/07/2025
17271	2025	All other reportable incidents	60789	Intracardiac pacemaker	CT1685	Cardiac pacemakers	E0604	Cardiac Perforation	A150302	Difficult or Delayed Separation		
							E0605	Cardiac Tamponade	A150302	Difficult or Delayed Separation		
							E0619	Pericardial Effusion	A150302	Difficult or Delayed Separation		
							E2321	Low Blood Pressure/ Hypotension	A150302	Difficult or Delayed Separation		
							F1203	Life Threatening Illness or Injury	A150302	Difficult or Delayed Separation		
					CT257	Implantable cardiac pacemakers	E0604	Cardiac Perforation	A150302	Difficult or Delayed Separation		
							E0605	Cardiac Tamponade	A150302	Difficult or Delayed Separation		
							E0619	Pericardial Effusion	A150302	Difficult or Delayed Separation		
							E2321	Low Blood Pressure/ Hypotension	A150302	Difficult or Delayed Separation		
							F1203	Life Threatening Illness or Injury	A150302	Difficult or Delayed Separation		
					CT2818	Catheterization/Cannulation/Fluid path devices	E0604	Cardiac Perforation	A150302	Difficult or Delayed Separation		
							E0605	Cardiac Tamponade	A150302	Difficult or Delayed Separation		
							E0619	Pericardial Effusion	A150302	Difficult or Delayed Separation		
							E2321	Low Blood Pressure/ Hypotension	A150302	Difficult or Delayed Separation		
							F1203	Life Threatening Illness or Injury	A150302	Difficult or Delayed Separation		
					CT2866	Implantable cardiac electrical stimulators	E0604	Cardiac Perforation	A150302	Difficult or Delayed Separation		
							E0605	Cardiac Tamponade	A150302	Difficult or Delayed Separation		
							E0619	Pericardial Effusion	A150302	Difficult or Delayed Separation		
							E2321	Low Blood Pressure/ Hypotension	A150302	Difficult or Delayed Separation		
							F1203	Life Threatening Illness or Injury	A150302	Difficult or Delayed Separation		
17272	2025	All other reportable incidents	57968	Cardiovascular device introducer, steerable	CT1684	Cardiac pacemakers and associated devices	E0604	Cardiac Perforation	A150302	Difficult or Delayed Separation		
							E0605	Cardiac Tamponade	A150302	Difficult or Delayed Separation		
							E0619	Pericardial Effusion	A150302	Difficult or Delayed Separation		
							E2321	Low Blood Pressure/ Hypotension	A150302	Difficult or Delayed Separation		
							F1203	Life Threatening Illness or Injury	A150302	Difficult or Delayed Separation		
17273	2025	All other reportable incidents	47270	Cardiac resynchronization therapy implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A070909	Oversensing		
									A071204	No Pacing		
							F23	Unexpected Medical Intervention	A070909	Oversensing		
									A071204	No Pacing		
17274	2025	All other reportable incidents	17882	Physiologic-monitoring defibrillation system	CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A071301	Failure to Deliver Shock/Stimulation		
							F27	Problem Identified During Non-Clinical Procedure	A071301	Failure to Deliver Shock/Stimulation		
17275	2025	Unanticipated serious deterioration in state of health	47265	Dual-chamber implantable pacemaker, rate-responsive	CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A0701	Capturing Problem		
							F12	Serious Injury/ Illness/ Impairment	A0701	Capturing Problem		
							F2307	Reprogramming of Device	A0701	Capturing Problem		
							F26	No Health Consequences or Impact	A0701	Capturing Problem		
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A0701	Capturing Problem		
							F12	Serious Injury/ Illness/ Impairment	A0701	Capturing Problem		
							F2307	Reprogramming of Device	A0701	Capturing Problem		
							F26	No Health Consequences or Impact	A0701	Capturing Problem		
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A0701	Capturing Problem		
							F12	Serious Injury/ Illness/ Impairment	A0701	Capturing Problem		
							F2307	Reprogramming of Device	A0701	Capturing Problem		
							F26	No Health Consequences or Impact	A0701	Capturing Problem		
17276	2025		47265	Dual-chamber implantable pacemaker, rate-responsive	CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	16/07/2025	16/07/2025
							F1901	Additional Surgery	A24	No Apparent Device Problem	16/07/2025	16/07/2025
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	16/07/2025	16/07/2025
							F1901	Additional Surgery	A24	No Apparent Device Problem	16/07/2025	16/07/2025
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	16/07/2025	16/07/2025
							F1901	Additional Surgery	A24	No Apparent Device Problem	16/07/2025	16/07/2025
17277	2025	All other reportable incidents	35852	Single-chamber implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A072201	High Impedance		
							F08	Hospitalization or Prolonged Hospitalization	A072201	High Impedance		
							F1905	Device Revision or Replacement	A072201	High Impedance		
17278	2025	All other reportable incidents	47265	Dual-chamber implantable pacemaker, rate-responsive	CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	23/12/2024	23/12/2024
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery	23/12/2024	23/12/2024
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	23/12/2024	23/12/2024
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery	23/12/2024	23/12/2024
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	23/12/2024	23/12/2024
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery	23/12/2024	23/12/2024
17279	2025	All other reportable incidents	47265	Dual-chamber implantable pacemaker, rate-responsive	CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A0203	Defective Device		
							F26	No Health Consequences or Impact	A0203	Defective Device		
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A0203	Defective Device		
							F26	No Health Consequences or Impact	A0203	Defective Device		
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A0203	Defective Device		
							F26	No Health Consequences or Impact	A0203	Defective Device		

17280	2025	All other reportable incidents	47265	Dual-chamber implantable pacemaker, rate-responsive	CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A0908	Incorrect, Inadequate or Imprecise Result or Readings	27/06/2025	27/06/2025
							F08	Hospitalization or Prolonged Hospitalization	A0908	Incorrect, Inadequate or Imprecise Result or Readings	27/06/2025	27/06/2025
							F1905	Device Revision or Replacement	A0908	Incorrect, Inadequate or Imprecise Result or Readings	27/06/2025	27/06/2025
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A0908	Incorrect, Inadequate or Imprecise Result or Readings	27/06/2025	27/06/2025
							F08	Hospitalization or Prolonged Hospitalization	A0908	Incorrect, Inadequate or Imprecise Result or Readings	27/06/2025	27/06/2025
							F1905	Device Revision or Replacement	A0908	Incorrect, Inadequate or Imprecise Result or Readings	27/06/2025	27/06/2025
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A0908	Incorrect, Inadequate or Imprecise Result or Readings	27/06/2025	27/06/2025
							F08	Hospitalization or Prolonged Hospitalization	A0908	Incorrect, Inadequate or Imprecise Result or Readings	27/06/2025	27/06/2025
							F1905	Device Revision or Replacement	A0908	Incorrect, Inadequate or Imprecise Result or Readings	27/06/2025	27/06/2025
17281	2025	All other reportable incidents	47265	Dual-chamber implantable pacemaker, rate-responsive	CT1685	Cardiac pacemakers	E060104	Bradycardia	A070909	Oversensing	16/07/2025	16/07/2025
									A0908	Incorrect, Inadequate or Imprecise Result or Readings	16/07/2025	16/07/2025
									A090801	Signal Artifact/Noise	16/07/2025	16/07/2025
							F08	Hospitalization or Prolonged Hospitalization	A070909	Oversensing	16/07/2025	16/07/2025
									A0908	Incorrect, Inadequate or Imprecise Result or Readings	16/07/2025	16/07/2025
									A090801	Signal Artifact/Noise	16/07/2025	16/07/2025
							F1905	Device Revision or Replacement	A070909	Oversensing	16/07/2025	16/07/2025
									A0908	Incorrect, Inadequate or Imprecise Result or Readings	16/07/2025	16/07/2025
									A090801	Signal Artifact/Noise	16/07/2025	16/07/2025
					CT257	Implantable cardiac pacemakers	E060104	Bradycardia	A070909	Oversensing	16/07/2025	16/07/2025
									A0908	Incorrect, Inadequate or Imprecise Result or Readings	16/07/2025	16/07/2025
									A090801	Signal Artifact/Noise	16/07/2025	16/07/2025
							F08	Hospitalization or Prolonged Hospitalization	A070909	Oversensing	16/07/2025	16/07/2025
									A0908	Incorrect, Inadequate or Imprecise Result or Readings	16/07/2025	16/07/2025
									A090801	Signal Artifact/Noise	16/07/2025	16/07/2025
							F1905	Device Revision or Replacement	A070909	Oversensing	16/07/2025	16/07/2025
									A0908	Incorrect, Inadequate or Imprecise Result or Readings	16/07/2025	16/07/2025
									A090801	Signal Artifact/Noise	16/07/2025	16/07/2025
					CT2866	Implantable cardiac electrical stimulators	E060104	Bradycardia	A070909	Oversensing	16/07/2025	16/07/2025
									A0908	Incorrect, Inadequate or Imprecise Result or Readings	16/07/2025	16/07/2025
									A090801	Signal Artifact/Noise	16/07/2025	16/07/2025
							F08	Hospitalization or Prolonged Hospitalization	A070909	Oversensing	16/07/2025	16/07/2025
									A0908	Incorrect, Inadequate or Imprecise Result or Readings	16/07/2025	16/07/2025
									A090801	Signal Artifact/Noise	16/07/2025	16/07/2025
							F1905	Device Revision or Replacement	A070909	Oversensing	16/07/2025	16/07/2025
									A0908	Incorrect, Inadequate or Imprecise Result or Readings	16/07/2025	16/07/2025
									A090801	Signal Artifact/Noise	16/07/2025	16/07/2025
17282	2025	All other reportable incidents	47265	Dual-chamber implantable pacemaker, rate-responsive			E060104	Bradycardia	A070504	Premature Discharge of Battery	16/07/2025	16/07/2025
									A0908	Incorrect, Inadequate or Imprecise Result or Readings	16/07/2025	16/07/2025
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	16/07/2025	16/07/2025
									A0908	Incorrect, Inadequate or Imprecise Result or Readings	16/07/2025	16/07/2025
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	16/07/2025	16/07/2025
									A0908	Incorrect, Inadequate or Imprecise Result or Readings	16/07/2025	16/07/2025
17283	2025	All other reportable incidents	37265	Dual-chamber implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A0712	Pacing Problem		
							F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A0712	Pacing Problem		
17284	2025	All other reportable incidents	47270	Cardiac resynchronization therapy implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
17285	2025	All other reportable incidents	47265	Dual-chamber implantable pacemaker, rate-responsive			E0119	Loss of Consciousness	A070909	Oversensing	21/07/2025	21/07/2025
									A0712	Pacing Problem	21/07/2025	21/07/2025
									A0908	Incorrect, Inadequate or Imprecise Result or Readings	21/07/2025	21/07/2025
									A090801	Signal Artifact/Noise	21/07/2025	21/07/2025
							E011903	Syncope/Fainting	A070909	Oversensing	21/07/2025	21/07/2025
									A0712	Pacing Problem	21/07/2025	21/07/2025
									A0908	Incorrect, Inadequate or Imprecise Result or Readings	21/07/2025	21/07/2025
									A090801	Signal Artifact/Noise	21/07/2025	21/07/2025
							E060101	Asystole	A070909	Oversensing	21/07/2025	21/07/2025
									A0712	Pacing Problem	21/07/2025	21/07/2025
									A0908	Incorrect, Inadequate or Imprecise Result or Readings	21/07/2025	21/07/2025
									A090801	Signal Artifact/Noise	21/07/2025	21/07/2025
							F08	Hospitalization or Prolonged Hospitalization	A070909	Oversensing	21/07/2025	21/07/2025
									A0712	Pacing Problem	21/07/2025	21/07/2025
									A0908	Incorrect, Inadequate or Imprecise Result or Readings	21/07/2025	21/07/2025
									A090801	Signal Artifact/Noise	21/07/2025	21/07/2025
							F1001	Absence of Treatment	A070909	Oversensing	21/07/2025	21/07/2025
									A0712	Pacing Problem	21/07/2025	21/07/2025
									A0908	Incorrect, Inadequate or Imprecise Result or Readings	21/07/2025	21/07/2025
									A090801	Signal Artifact/Noise	21/07/2025	21/07/2025
							F1905	Device Revision or Replacement	A070909	Oversensing	21/07/2025	21/07/2025
									A0712	Pacing Problem	21/07/2025	21/07/2025

									A0908	Incorrect, Inadequate or Imprecise Result or Readings	21/07/2025	21/07/2025
									A090801	Signal Artifact/Noise	21/07/2025	21/07/2025
17286	2025	All other reportable incidents	47265	Dual-chamber implantable pacemaker, rate-responsive	CT1685	Cardiac pacemakers	E2312	Fatigue	A24	No Apparent Device Problem		
							E2402	Appropriate Clinical Signs, Symptoms and Conditions Term/Code not Available	A24	No Apparent Device Problem		
							F12	Serious Injury/ Illness/ Impairment	A24	No Apparent Device Problem		
							F2307	Reprogramming of Device	A24	No Apparent Device Problem		
							F28	Appropriate Health Impact Term/Code not Available	A24	No Apparent Device Problem		
					CT257	Implantable cardiac pacemakers	E2312	Fatigue	A24	No Apparent Device Problem		
							E2402	Appropriate Clinical Signs, Symptoms and Conditions Term/Code not Available	A24	No Apparent Device Problem		
							F12	Serious Injury/ Illness/ Impairment	A24	No Apparent Device Problem		
							F2307	Reprogramming of Device	A24	No Apparent Device Problem		
							F28	Appropriate Health Impact Term/Code not Available	A24	No Apparent Device Problem		
					CT2866	Implantable cardiac electrical stimulators	E2312	Fatigue	A24	No Apparent Device Problem		
							E2402	Appropriate Clinical Signs, Symptoms and Conditions Term/Code not Available	A24	No Apparent Device Problem		
							F12	Serious Injury/ Illness/ Impairment	A24	No Apparent Device Problem		
							F2307	Reprogramming of Device	A24	No Apparent Device Problem		
							F28	Appropriate Health Impact Term/Code not Available	A24	No Apparent Device Problem		
17287	2025	All other reportable incidents	47265	Dual-chamber implantable pacemaker, rate-responsive	CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A1606	Premature Indicator Activation		
							F26	No Health Consequences or Impact	A1606	Premature Indicator Activation		
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A1606	Premature Indicator Activation		
							F26	No Health Consequences or Impact	A1606	Premature Indicator Activation		
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A1606	Premature Indicator Activation		
							F26	No Health Consequences or Impact	A1606	Premature Indicator Activation		
17288	2025	All other reportable incidents	47265	Dual-chamber implantable pacemaker, rate-responsive			E2403	No Clinical Signs, Symptoms or Conditions	A0908	Incorrect, Inadequate or Imprecise Result or Readings	10/08/2025	10/08/2025
							F08	Hospitalization or Prolonged Hospitalization	A0908	Incorrect, Inadequate or Imprecise Result or Readings	10/08/2025	10/08/2025
							F1905	Device Revision or Replacement	A0908	Incorrect, Inadequate or Imprecise Result or Readings	10/08/2025	10/08/2025
17289	2025	All other reportable incidents	47270	Cardiac resynchronization therapy implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A040101	Fracture		
									A072201	High Impedance		
							F26	No Health Consequences or Impact	A040101	Fracture		
									A072201	High Impedance		
17290	2025	All other reportable incidents	47270	Cardiac resynchronization therapy implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E210401	Shock from Patient Lead(s)	A071302	Inappropriate/Inadequate Shock/Stimulation		
							F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A071302	Inappropriate/Inadequate Shock/Stimulation		
17291	2025	All other reportable incidents	47270	Cardiac resynchronization therapy implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E210401	Shock from Patient Lead(s)	A070909	Oversensing		
									A071302	Inappropriate/Inadequate Shock/Stimulation		
									A090801	Signal Artifact/Noise		
									A1301	Failure to Read Input Signal		
							F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A070909	Oversensing		
									A071302	Inappropriate/Inadequate Shock/Stimulation		
									A090801	Signal Artifact/Noise		
									A1301	Failure to Read Input Signal		
17292	2025	All other reportable incidents	47265	Dual-chamber implantable pacemaker, rate-responsive	CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	30/04/2025	30/04/2025
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	30/04/2025	30/04/2025
							F1903	Device Explantation	A070504	Premature Discharge of Battery	30/04/2025	30/04/2025
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	30/04/2025	30/04/2025
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	30/04/2025	30/04/2025
							F1903	Device Explantation	A070504	Premature Discharge of Battery	30/04/2025	30/04/2025
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	30/04/2025	30/04/2025
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	30/04/2025	30/04/2025
							F1903	Device Explantation	A070504	Premature Discharge of Battery	30/04/2025	30/04/2025
17293	2025	All other reportable incidents	37265	Dual-chamber implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A070102	High Capture Threshold		
									A072201	High Impedance		
									A090801	Signal Artifact/Noise		
							F26	No Health Consequences or Impact	A070102	High Capture Threshold		
									A072201	High Impedance		
									A090801	Signal Artifact/Noise		
17294	2025	All other reportable incidents	47270	Cardiac resynchronization therapy implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
17295	2025	All other reportable incidents	47263	Cardiac resynchronization therapy implantable pacemaker	CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A05	Mechanical Problem	21/07/2025	21/07/2025
							F26	No Health Consequences or Impact	A05	Mechanical Problem	21/07/2025	21/07/2025
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A05	Mechanical Problem	21/07/2025	21/07/2025
							F26	No Health Consequences or Impact	A05	Mechanical Problem	21/07/2025	21/07/2025
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A05	Mechanical Problem	21/07/2025	21/07/2025
							F26	No Health Consequences or Impact	A05	Mechanical Problem	21/07/2025	21/07/2025
17296	2025	All other reportable incidents	60956	Pacing/defibrillation lead tunneller	CT1684	Cardiac pacemakers and associated devices	E210401	Shock from Patient Lead(s)	A070909	Oversensing		
									A071302	Inappropriate/Inadequate Shock/Stimulation		
							F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A070909	Oversensing		
									A071302	Inappropriate/Inadequate Shock/Stimulation		
17297	2025	All other reportable incidents	37265	Dual-chamber implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A070102	High Capture Threshold		
									A072201	High Impedance		
									A1206	Misconnection		
							F26	No Health Consequences or Impact	A070102	High Capture Threshold		
									A072201	High Impedance		
									A1206	Misconnection		
17298	2024	Death	17882	Physiologic-monitoring defibrillation system	CT1684	Cardiac pacemakers and associated devices	E2401	Insufficient Information	A23	Use of Device Problem		
							F02	Death	A23	Use of Device Problem		
17299	2025	All other reportable incidents	47205	Cardiac pulse generator programmer	CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A1301	Failure to Read Input Signal		
							F08	Hospitalization or Prolonged Hospitalization	A1301	Failure to Read Input Signal		
							F2203	Imaging Required	A1301	Failure to Read Input Signal		
17300	2025	All other reportable incidents	47270	Cardiac resynchronization therapy implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A070909	Oversensing		
									A090801	Signal Artifact/Noise		
							F26	No Health Consequences or Impact	A070909	Oversensing		
									A090801	Signal Artifact/Noise		

17301	2025	All other reportable incidents	47270	Cardiac resynchronization therapy implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E210401	Shock from Patient Lead(s)	A070909	Oversensing		
									A071302	Inappropriate/Inadequate Shock/Stimulation		
									A1301	Failure to Read Input Signal		
							F08	Hospitalization or Prolonged Hospitalization	A070909	Oversensing		
									A071302	Inappropriate/Inadequate Shock/Stimulation		
									A1301	Failure to Read Input Signal		
							F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A070909	Oversensing		
									A071302	Inappropriate/Inadequate Shock/Stimulation		
									A1301	Failure to Read Input Signal		
							F2203	Imaging Required	A070909	Oversensing		
									A071302	Inappropriate/Inadequate Shock/Stimulation		
									A1301	Failure to Read Input Signal		
							F23	Unexpected Medical Intervention	A070909	Oversensing		
									A071302	Inappropriate/Inadequate Shock/Stimulation		
									A1301	Failure to Read Input Signal		
17302	2025	All other reportable incidents	17882	Physiologic-monitoring defibrillation system	CT1684	Cardiac pacemakers and associated devices	E0602	Cardiac Arrest	A071301	Failure to Deliver Shock/Stimulation		
							F02	Death	A071301	Failure to Deliver Shock/Stimulation		
17303	2025		60956	Pacing/defibrillation lead tunneller	CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
									A160602	Premature End-of-Life Indicator		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
									A160602	Premature End-of-Life Indicator		
17304	2025	All other reportable incidents	35852	Single-chamber implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E210401	Shock from Patient Lead(s)	A070909	Oversensing		
									A071302	Inappropriate/Inadequate Shock/Stimulation		
							F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A070909	Oversensing		
									A071302	Inappropriate/Inadequate Shock/Stimulation		
17305	2025	All other reportable incidents	47270	Cardiac resynchronization therapy implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A0709	Device Sensing Problem		
									A0712	Pacing Problem		
							F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A0709	Device Sensing Problem		
									A0712	Pacing Problem		
							F12	Serious Injury/ Illness/ Impairment	A0709	Device Sensing Problem		
									A0712	Pacing Problem		
							F2307	Reprogramming of Device	A0709	Device Sensing Problem		
									A0712	Pacing Problem		
17306	2025	All other reportable incidents	35852	Single-chamber implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E060109	Tachycardia	A0709	Device Sensing Problem		
									A071301	Failure to Deliver Shock/Stimulation		
							F1001	Absence of Treatment	A0709	Device Sensing Problem		
									A071301	Failure to Deliver Shock/Stimulation		
							F1203	Life Threatening Illness or Injury	A0709	Device Sensing Problem		
									A071301	Failure to Deliver Shock/Stimulation		
17307	2025	All other reportable incidents	61126	Implantable pulse generator mesh bag, bioabsorbable	CT1684	Cardiac pacemakers and associated devices	E0306	Sepsis	A24	No Apparent Device Problem	09/07/2025	09/07/2025
							E1906	Unspecified Infection	A24	No Apparent Device Problem	09/07/2025	09/07/2025
							F08	Hospitalization or Prolonged Hospitalization	A24	No Apparent Device Problem	09/07/2025	09/07/2025
							F1203	Life Threatening Illness or Injury	A24	No Apparent Device Problem	09/07/2025	09/07/2025
							F1903	Device Explantation	A24	No Apparent Device Problem	09/07/2025	09/07/2025
17308	2025	Unanticipated serious deterioration in state of health	37265	Dual-chamber implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E0306	Sepsis	A24	No Apparent Device Problem	09/07/2025	09/07/2025
							E1906	Unspecified Infection	A24	No Apparent Device Problem	09/07/2025	09/07/2025
							F08	Hospitalization or Prolonged Hospitalization	A24	No Apparent Device Problem	09/07/2025	09/07/2025
							F1203	Life Threatening Illness or Injury	A24	No Apparent Device Problem	09/07/2025	09/07/2025
							F1903	Device Explantation	A24	No Apparent Device Problem	09/07/2025	09/07/2025
17309	2025	All other reportable incidents	47263	Cardiac resynchronization therapy implantable pacemaker	CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A070101	Failure to Capture		
							F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A070101	Failure to Capture		
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A070101	Failure to Capture		
							F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A070101	Failure to Capture		
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A070101	Failure to Capture		
							F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A070101	Failure to Capture		
17310	2025	All other reportable incidents	47265	Dual-chamber implantable pacemaker, rate-responsive	CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A070101	Failure to Capture		
									A070909	Oversensing		
							F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A070101	Failure to Capture		
									A070909	Oversensing		
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A070101	Failure to Capture		
									A070909	Oversensing		
							F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A070101	Failure to Capture		
									A070909	Oversensing		
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A070101	Failure to Capture		
									A070909	Oversensing		
							F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A070101	Failure to Capture		
									A070909	Oversensing		
17311	2025	All other reportable incidents	37265	Dual-chamber implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A0709	Device Sensing Problem		
							F12	Serious Injury/ Illness/ Impairment	A0709	Device Sensing Problem		
							F2307	Reprogramming of Device	A0709	Device Sensing Problem		
17312	2025	All other reportable incidents	47270	Cardiac resynchronization therapy implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A070910	Undersensing		
									A0712	Pacing Problem		
							F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A070910	Undersensing		
									A0712	Pacing Problem		
17313	2025	All other reportable incidents	35852	Single-chamber implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E210401	Shock from Patient Lead(s)	A0709	Device Sensing Problem		
									A071302	Inappropriate/Inadequate Shock/Stimulation		
							F08	Hospitalization or Prolonged Hospitalization	A0709	Device Sensing Problem		
									A071302	Inappropriate/Inadequate Shock/Stimulation		
							F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A0709	Device Sensing Problem		
									A071302	Inappropriate/Inadequate Shock/Stimulation		
							F1203	Life Threatening Illness or Injury	A0709	Device Sensing Problem		
									A071302	Inappropriate/Inadequate Shock/Stimulation		
							F2307	Reprogramming of Device	A0709	Device Sensing Problem		
									A071302	Inappropriate/Inadequate Shock/Stimulation		
17314	2025	All other reportable incidents	47265	Dual-chamber implantable pacemaker, rate-responsive	CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A0407	Material Discolored	05/08/2025	05/08/2025
									A0701	Capturing Problem	05/08/2025	05/08/2025
									A070504	Premature Discharge of Battery	05/08/2025	05/08/2025
							F1901	Additional Surgery	A0407	Material Discolored	05/08/2025	05/08/2025

									A0701	Capturing Problem	05/08/2025	05/08/2025
									A070504	Premature Discharge of Battery	05/08/2025	05/08/2025
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A0407	Material Discolored	05/08/2025	05/08/2025
									A0701	Capturing Problem	05/08/2025	05/08/2025
									A070504	Premature Discharge of Battery	05/08/2025	05/08/2025
							F1901	Additional Surgery	A0407	Material Discolored	05/08/2025	05/08/2025
									A0701	Capturing Problem	05/08/2025	05/08/2025
									A070504	Premature Discharge of Battery	05/08/2025	05/08/2025
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A0407	Material Discolored	05/08/2025	05/08/2025
									A0701	Capturing Problem	05/08/2025	05/08/2025
									A070504	Premature Discharge of Battery	05/08/2025	05/08/2025
							F1901	Additional Surgery	A0407	Material Discolored	05/08/2025	05/08/2025
									A0701	Capturing Problem	05/08/2025	05/08/2025
									A070504	Premature Discharge of Battery	05/08/2025	05/08/2025
17315	2025	All other reportable incidents	47267	Single-chamber implantable pacemaker, rate-responsive	CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	08/07/2025	08/07/2025
							F1901	Additional Surgery	A24	No Apparent Device Problem	08/07/2025	08/07/2025
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	08/07/2025	08/07/2025
							F1901	Additional Surgery	A24	No Apparent Device Problem	08/07/2025	08/07/2025
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	08/07/2025	08/07/2025
							F1901	Additional Surgery	A24	No Apparent Device Problem	08/07/2025	08/07/2025
17316	2025	All other reportable incidents	47270	Cardiac resynchronization therapy implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A0908	Incorrect, Inadequate or Imprecise Result or Readings		
									A1107	Data Problem		
							F26	No Health Consequences or Impact	A0908	Incorrect, Inadequate or Imprecise Result or Readings		
									A1107	Data Problem		
17317	2025	All other reportable incidents	35852	Single-chamber implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E2104	Electric Shock	A071201	Failure to Convert Rhythm	29/07/2025	29/07/2025
							F1908	Prolonged Surgery	A071201	Failure to Convert Rhythm	29/07/2025	29/07/2025
							F2306	Resuscitation	A071201	Failure to Convert Rhythm	29/07/2025	29/07/2025
17318	2025	All other reportable incidents	35852	Single-chamber implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A070910	Undersensing		
									A071302	Inappropriate/Inadequate Shock/Stimulation		
							F12	Serious Injury/ Illness/ Impairment	A070910	Undersensing		
									A071302	Inappropriate/Inadequate Shock/Stimulation		
17319	2025	All other reportable incidents	47205	Cardiac pulse generator programmer	CT1684	Cardiac pacemakers and associated devices	E210401	Shock from Patient Lead(s)	A070909	Oversensing		
									A071302	Inappropriate/Inadequate Shock/Stimulation		
									A090801	Signal Artifact/Noise		
							F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A070909	Oversensing		
									A071302	Inappropriate/Inadequate Shock/Stimulation		
									A090801	Signal Artifact/Noise		
							F2203	Imaging Required	A070909	Oversensing		
									A071302	Inappropriate/Inadequate Shock/Stimulation		
									A090801	Signal Artifact/Noise		
							F23	Unexpected Medical Intervention	A070909	Oversensing		
									A071302	Inappropriate/Inadequate Shock/Stimulation		
									A090801	Signal Artifact/Noise		
17320	2025	All other reportable incidents	47263	Cardiac resynchronization therapy implantable pacemaker	CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A072001	Electromagnetic Interference		
							F26	No Health Consequences or Impact	A072001	Electromagnetic Interference		
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A072001	Electromagnetic Interference		
							F26	No Health Consequences or Impact	A072001	Electromagnetic Interference		
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A072001	Electromagnetic Interference		
							F26	No Health Consequences or Impact	A072001	Electromagnetic Interference		
17321	2025	All other reportable incidents	47270	Cardiac resynchronization therapy implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E060110	Ventricular Fibrillation	A070909	Oversensing		
									A070910	Undersensing		
									A071202	Inaccurate Synchronization		
									A0713	Defibrillation/Stimulation Problem		
									A071302	Inappropriate/Inadequate Shock/Stimulation		
									A1301	Failure to Read Input Signal		
							E210401	Shock from Patient Lead(s)	A070909	Oversensing		
									A070910	Undersensing		
									A071202	Inaccurate Synchronization		
									A0713	Defibrillation/Stimulation Problem		
									A071302	Inappropriate/Inadequate Shock/Stimulation		
									A1301	Failure to Read Input Signal		
							F05	Delay to Treatment/ Therapy	A070909	Oversensing		
									A070910	Undersensing		
									A071202	Inaccurate Synchronization		
									A0713	Defibrillation/Stimulation Problem		
									A071302	Inappropriate/Inadequate Shock/Stimulation		
									A1301	Failure to Read Input Signal		
							F08	Hospitalization or Prolonged Hospitalization	A070909	Oversensing		
									A070910	Undersensing		
									A071202	Inaccurate Synchronization		
									A0713	Defibrillation/Stimulation Problem		
									A071302	Inappropriate/Inadequate Shock/Stimulation		
									A1301	Failure to Read Input Signal		
							F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A070909	Oversensing		
									A070910	Undersensing		
									A071202	Inaccurate Synchronization		
									A0713	Defibrillation/Stimulation Problem		
									A071302	Inappropriate/Inadequate Shock/Stimulation		
									A1301	Failure to Read Input Signal		
							F1203	Life Threatening Illness or Injury	A070909	Oversensing		
									A070910	Undersensing		
									A071202	Inaccurate Synchronization		
									A0713	Defibrillation/Stimulation Problem		
									A071302	Inappropriate/Inadequate Shock/Stimulation		
									A1301	Failure to Read Input Signal		
							F15	Recognized Device or Procedural Complication	A070909	Oversensing		

									A070910	Undersensing		
									A071202	Inaccurate Synchronization		
									A0713	Defibrillation/Stimulation Problem		
									A071302	Inappropriate/Inadequate Shock/Stimulation		
									A1301	Failure to Read Input Signal		
17322	2025	All other reportable incidents	47265	Dual-chamber implantable pacemaker, rate-responsive			E060101	Asystole	A070909	Oversensing	31/07/2025	31/07/2025
									A0908	Incorrect, Inadequate or Imprecise Result or Readings	31/07/2025	31/07/2025
							F08	Hospitalization or Prolonged Hospitalization	A070909	Oversensing	31/07/2025	31/07/2025
									A0908	Incorrect, Inadequate or Imprecise Result or Readings	31/07/2025	31/07/2025
							F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A070909	Oversensing	31/07/2025	31/07/2025
									A0908	Incorrect, Inadequate or Imprecise Result or Readings	31/07/2025	31/07/2025
							F1905	Device Revision or Replacement	A070909	Oversensing	31/07/2025	31/07/2025
									A0908	Incorrect, Inadequate or Imprecise Result or Readings	31/07/2025	31/07/2025
17323	2025	All other reportable incidents	47270	Cardiac resynchronization therapy implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A072201	High Impedance		
							F26	No Health Consequences or Impact	A072201	High Impedance		
17324	2025	All other reportable incidents	47270	Cardiac resynchronization therapy implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A070909	Oversensing		
									A090801	Signal Artifact/Noise		
							F26	No Health Consequences or Impact	A070909	Oversensing		
									A090801	Signal Artifact/Noise		
17325	2025	All other reportable incidents	47265	Dual-chamber implantable pacemaker, rate-responsive	CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A071102	Failure to Interrogate	01/07/2025	31/07/2025
									A071204	No Pacing	01/07/2025	31/07/2025
							F1901	Additional Surgery	A071102	Failure to Interrogate	01/07/2025	31/07/2025
									A071204	No Pacing	01/07/2025	31/07/2025
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A071102	Failure to Interrogate	01/07/2025	31/07/2025
									A071204	No Pacing	01/07/2025	31/07/2025
							F1901	Additional Surgery	A071102	Failure to Interrogate	01/07/2025	31/07/2025
									A071204	No Pacing	01/07/2025	31/07/2025
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A071102	Failure to Interrogate	01/07/2025	31/07/2025
									A071204	No Pacing	01/07/2025	31/07/2025
							F1901	Additional Surgery	A071102	Failure to Interrogate	01/07/2025	31/07/2025
									A071204	No Pacing	01/07/2025	31/07/2025
17326	2025	All other reportable incidents	35852	Single-chamber implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E060109	Tachycardia	A0709	Device Sensing Problem		
									A071301	Failure to Deliver Shock/Stimulation		
							F1001	Absence of Treatment	A0709	Device Sensing Problem		
									A071301	Failure to Deliver Shock/Stimulation		
							F1203	Life Threatening Illness or Injury	A0709	Device Sensing Problem		
									A071301	Failure to Deliver Shock/Stimulation		
17327	2025	All other reportable incidents	17882	Physiologic-monitoring defibrillation system	CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A1603	Failure of Device to Self-Test		
							F27	Problem Identified During Non-Clinical Procedure	A1603	Failure of Device to Self-Test		
17328	2025	All other reportable incidents	47265	Dual-chamber implantable pacemaker, rate-responsive			E2403	No Clinical Signs, Symptoms or Conditions	A0908	Incorrect, Inadequate or Imprecise Result or Readings	03/09/2025	03/09/2025
							F08	Hospitalization or Prolonged Hospitalization	A0908	Incorrect, Inadequate or Imprecise Result or Readings	03/09/2025	03/09/2025
							F1905	Device Revision or Replacement	A0908	Incorrect, Inadequate or Imprecise Result or Readings	03/09/2025	03/09/2025
17329	2025	All other reportable incidents	47267	Single-chamber implantable pacemaker, rate-responsive	CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem		
							F26	No Health Consequences or Impact	A0705	Battery Problem		
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem		
							F26	No Health Consequences or Impact	A0705	Battery Problem		
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem		
							F26	No Health Consequences or Impact	A0705	Battery Problem		
17330	2025	All other reportable incidents	47265	Dual-chamber implantable pacemaker, rate-responsive	CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A070102	High Capture Threshold		
									A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070102	High Capture Threshold		
									A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A070102	High Capture Threshold		
									A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070102	High Capture Threshold		
									A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A070102	High Capture Threshold		
									A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070102	High Capture Threshold		
									A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
17331	2025	All other reportable incidents	47270	Cardiac resynchronization therapy implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A072201	High Impedance		
							F26	No Health Consequences or Impact	A072201	High Impedance		
17332	2025	All other reportable incidents	60956	Pacing/defibrillation lead tunneller	CT1684	Cardiac pacemakers and associated devices	E210401	Shock from Patient Lead(s)	A040101	Fracture		
									A070909	Oversensing		
									A071302	Inappropriate/Inadequate Shock/Stimulation		
									A090801	Signal Artifact/Noise		
							F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A040101	Fracture		
									A070909	Oversensing		
									A071302	Inappropriate/Inadequate Shock/Stimulation		
									A090801	Signal Artifact/Noise		
							F2203	Imaging Required	A040101	Fracture		
									A070909	Oversensing		
									A071302	Inappropriate/Inadequate Shock/Stimulation		
									A090801	Signal Artifact/Noise		
							F23	Unexpected Medical Intervention	A040101	Fracture		
									A070909	Oversensing		

									A071302	Inappropriate/Inadequate Shock/Stimulation		
									A090801	Signal Artifact/Noise		
17333	2025	All other reportable incidents	47265	Dual-chamber implantable pacemaker, rate-responsive	CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A070910	Undersensing		
							F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A070910	Undersensing		
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A070910	Undersensing		
							F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A070910	Undersensing		
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A070910	Undersensing		
							F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A070910	Undersensing		
17334	2025	All other reportable incidents	37265	Dual-chamber implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A070909	Oversensing	24/06/2025	24/06/2025
									A0712	Pacing Problem	24/06/2025	24/06/2025
									A072201	High Impedance	24/06/2025	24/06/2025
									A090801	Signal Artifact/Noise	24/06/2025	24/06/2025
									A1301	Failure to Read Input Signal	24/06/2025	24/06/2025
							F08	Hospitalization or Prolonged Hospitalization	A070909	Oversensing	24/06/2025	24/06/2025
									A0712	Pacing Problem	24/06/2025	24/06/2025
									A072201	High Impedance	24/06/2025	24/06/2025
									A090801	Signal Artifact/Noise	24/06/2025	24/06/2025
									A1301	Failure to Read Input Signal	24/06/2025	24/06/2025
							F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A070909	Oversensing	24/06/2025	24/06/2025
									A0712	Pacing Problem	24/06/2025	24/06/2025
									A072201	High Impedance	24/06/2025	24/06/2025
									A090801	Signal Artifact/Noise	24/06/2025	24/06/2025
									A1301	Failure to Read Input Signal	24/06/2025	24/06/2025
							F1905	Device Revision or Replacement	A070909	Oversensing	24/06/2025	24/06/2025
									A0712	Pacing Problem	24/06/2025	24/06/2025
									A072201	High Impedance	24/06/2025	24/06/2025
									A090801	Signal Artifact/Noise	24/06/2025	24/06/2025
									A1301	Failure to Read Input Signal	24/06/2025	24/06/2025
							F23	Unexpected Medical Intervention	A070909	Oversensing	24/06/2025	24/06/2025
									A0712	Pacing Problem	24/06/2025	24/06/2025
									A072201	High Impedance	24/06/2025	24/06/2025
									A090801	Signal Artifact/Noise	24/06/2025	24/06/2025
									A1301	Failure to Read Input Signal	24/06/2025	24/06/2025
17335	2025	All other reportable incidents	35852	Single-chamber implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E210401	Shock from Patient Lead(s)	A0709	Device Sensing Problem		
									A0712	Pacing Problem		
									A071302	Inappropriate/Inadequate Shock/Stimulation		
							F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A0709	Device Sensing Problem		
									A0712	Pacing Problem		
									A071302	Inappropriate/Inadequate Shock/Stimulation		
							F1203	Life Threatening Illness or Injury	A0709	Device Sensing Problem		
									A0712	Pacing Problem		
									A071302	Inappropriate/Inadequate Shock/Stimulation		
							F2307	Reprogramming of Device	A0709	Device Sensing Problem		
									A0712	Pacing Problem		
									A071302	Inappropriate/Inadequate Shock/Stimulation		
17336	2025	All other reportable incidents	37265	Dual-chamber implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A071102	Failure to Interrogate		
							F26	No Health Consequences or Impact	A071102	Failure to Interrogate		
17337	2025	All other reportable incidents	47265	Dual-chamber implantable pacemaker, rate-responsive			E2403	No Clinical Signs, Symptoms or Conditions	A0908	Incorrect, Inadequate or Imprecise Result or Readings	06/08/2025	06/08/2025
							F08	Hospitalization or Prolonged Hospitalization	A0908	Incorrect, Inadequate or Imprecise Result or Readings	06/08/2025	06/08/2025
							F1905	Device Revision or Replacement	A0908	Incorrect, Inadequate or Imprecise Result or Readings	06/08/2025	06/08/2025
17338	2025	All other reportable incidents	47270	Cardiac resynchronization therapy implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
17339	2025	All other reportable incidents	47265	Dual-chamber implantable pacemaker, rate-responsive			E2403	No Clinical Signs, Symptoms or Conditions	A0908	Incorrect, Inadequate or Imprecise Result or Readings	08/10/2025	08/10/2025
							F08	Hospitalization or Prolonged Hospitalization	A0908	Incorrect, Inadequate or Imprecise Result or Readings	08/10/2025	08/10/2025
							F1905	Device Revision or Replacement	A0908	Incorrect, Inadequate or Imprecise Result or Readings	08/10/2025	08/10/2025
17340	2025	All other reportable incidents	47270	Cardiac resynchronization therapy implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A070101	Failure to Capture		
							F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A070101	Failure to Capture		
							F2203	Imaging Required	A070101	Failure to Capture		
17341	2025		47265	Dual-chamber implantable pacemaker, rate-responsive	CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A0508	Noise, Audible	14/08/2025	14/08/2025
									A070909	Oversensing	14/08/2025	14/08/2025
									A0712	Pacing Problem	14/08/2025	14/08/2025
							F1901	Additional Surgery	A0508	Noise, Audible	14/08/2025	14/08/2025
									A070909	Oversensing	14/08/2025	14/08/2025
									A0712	Pacing Problem	14/08/2025	14/08/2025
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A0508	Noise, Audible	14/08/2025	14/08/2025
									A070909	Oversensing	14/08/2025	14/08/2025
									A0712	Pacing Problem	14/08/2025	14/08/2025
							F1901	Additional Surgery	A0508	Noise, Audible	14/08/2025	14/08/2025
									A070909	Oversensing	14/08/2025	14/08/2025
									A0712	Pacing Problem	14/08/2025	14/08/2025
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A0508	Noise, Audible	14/08/2025	14/08/2025
									A070909	Oversensing	14/08/2025	14/08/2025
									A0712	Pacing Problem	14/08/2025	14/08/2025
							F1901	Additional Surgery	A0508	Noise, Audible	14/08/2025	14/08/2025
									A070909	Oversensing	14/08/2025	14/08/2025
									A0712	Pacing Problem	14/08/2025	14/08/2025
17342	2025	All other reportable incidents	47267	Single-chamber implantable pacemaker, rate-responsive	CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A070101	Failure to Capture		
							F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A070101	Failure to Capture		
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A070101	Failure to Capture		
							F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A070101	Failure to Capture		
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A070101	Failure to Capture		
							F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A070101	Failure to Capture		
17343	2025	All other reportable incidents	47270	Cardiac resynchronization therapy implantable	CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A0709	Device Sensing Problem		

				defibrillator					A071302	Inappropriate/Inadequate Shock/Stimulation		
							F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A0709	Device Sensing Problem		
									A071302	Inappropriate/Inadequate Shock/Stimulation		
							F1203	Life Threatening Illness or Injury	A0709	Device Sensing Problem		
									A071302	Inappropriate/Inadequate Shock/Stimulation		
							F2307	Reprogramming of Device	A0709	Device Sensing Problem		
									A071302	Inappropriate/Inadequate Shock/Stimulation		
17344	2025	All other reportable incidents	47265	Dual-chamber implantable pacemaker, rate-responsive	CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A05	Mechanical Problem	05/08/2025	05/08/2025
							F26	No Health Consequences or Impact	A05	Mechanical Problem	05/08/2025	05/08/2025
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A05	Mechanical Problem	05/08/2025	05/08/2025
							F26	No Health Consequences or Impact	A05	Mechanical Problem	05/08/2025	05/08/2025
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A05	Mechanical Problem	05/08/2025	05/08/2025
							F26	No Health Consequences or Impact	A05	Mechanical Problem	05/08/2025	05/08/2025
17345	2025	All other reportable incidents	17882	Physiologic-monitoring defibrillation system	CT1684	Cardiac pacemakers and associated devices	E0602	Cardiac Arrest	A071301	Failure to Deliver Shock/Stimulation		
							F02	Death	A071301	Failure to Deliver Shock/Stimulation		
17346	2025		47267	Single-chamber implantable pacemaker, rate-responsive	CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A05	Mechanical Problem		
									A0501	Detachment of Device or Device Component		
							F26	No Health Consequences or Impact	A05	Mechanical Problem		
									A0501	Detachment of Device or Device Component		
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A05	Mechanical Problem		
									A0501	Detachment of Device or Device Component		
							F26	No Health Consequences or Impact	A05	Mechanical Problem		
									A0501	Detachment of Device or Device Component		
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A05	Mechanical Problem		
									A0501	Detachment of Device or Device Component		
							F26	No Health Consequences or Impact	A05	Mechanical Problem		
									A0501	Detachment of Device or Device Component		
17347	2025	All other reportable incidents	47267	Single-chamber implantable pacemaker, rate-responsive			E2403	No Clinical Signs, Symptoms or Conditions	A0908	Incorrect, Inadequate or Imprecise Result or Readings	23/07/2025	23/07/2025
							F08	Hospitalization or Prolonged Hospitalization	A0908	Incorrect, Inadequate or Imprecise Result or Readings	23/07/2025	23/07/2025
							F1905	Device Revision or Replacement	A0908	Incorrect, Inadequate or Imprecise Result or Readings	23/07/2025	23/07/2025
17348	2025	All other reportable incidents	47270	Cardiac resynchronization therapy implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A072201	High Impedance		
							F26	No Health Consequences or Impact	A072201	High Impedance		
17349	2025	All other reportable incidents	37265	Dual-chamber implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
17350	2025	All other reportable incidents	47270	Cardiac resynchronization therapy implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A070909	Oversensing		
									A090801	Signal Artifact/Noise		
							F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A070909	Oversensing		
									A090801	Signal Artifact/Noise		
17351	2025	All other reportable incidents	37265	Dual-chamber implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A070102	High Capture Threshold		
									A072201	High Impedance		
							F2203	Imaging Required	A070102	High Capture Threshold		
									A072201	High Impedance		
17352	2025	All other reportable incidents	36106	Implantable pacemaker analyser, intraoperative	CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A0712	Pacing Problem		
							F26	No Health Consequences or Impact	A0712	Pacing Problem		
17353	2025	All other reportable incidents	36106	Implantable pacemaker analyser, intraoperative	CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A11	Computer Software Problem		
							F26	No Health Consequences or Impact	A11	Computer Software Problem		
17354	2025	All other reportable incidents	36106	Implantable pacemaker analyser, intraoperative	CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A11	Computer Software Problem		
							F26	No Health Consequences or Impact	A11	Computer Software Problem		
17355	2025	All other reportable incidents	47265	Dual-chamber implantable pacemaker, rate-responsive	CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A070909	Oversensing		
									A070910	Undersensing		
									A072001	Electromagnetic Interference		
									A090801	Signal Artifact/Noise		
							F26	No Health Consequences or Impact	A070909	Oversensing		
									A070910	Undersensing		
									A072001	Electromagnetic Interference		
									A090801	Signal Artifact/Noise		
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A070909	Oversensing		
									A070910	Undersensing		
									A072001	Electromagnetic Interference		
									A090801	Signal Artifact/Noise		
							F26	No Health Consequences or Impact	A070909	Oversensing		
									A070910	Undersensing		
									A072001	Electromagnetic Interference		
									A090801	Signal Artifact/Noise		
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A070909	Oversensing		
									A070910	Undersensing		
									A072001	Electromagnetic Interference		
									A090801	Signal Artifact/Noise		
							F26	No Health Consequences or Impact	A070909	Oversensing		
									A070910	Undersensing		
									A072001	Electromagnetic Interference		
									A090801	Signal Artifact/Noise		
17356	2025	All other reportable incidents	63188	Left-ventricular wireless pacing system sensor/transmitter	CT145	Cardiovascular devices	E0623	Unspecified Heart Problem	A0705	Battery Problem		
							F24	Insufficient Health Impact Information	A0705	Battery Problem		
					CT1577	Wireless monitoring system transmitters	E0623	Unspecified Heart Problem	A0705	Battery Problem		
							F24	Insufficient Health Impact Information	A0705	Battery Problem		
					CT1684	Cardiac pacemakers and associated devices	E0623	Unspecified Heart Problem	A0705	Battery Problem		
							F24	Insufficient Health Impact Information	A0705	Battery Problem		
					CT2727	Active implant devices/ancillaries	E0623	Unspecified Heart Problem	A0705	Battery Problem		
							F24	Insufficient Health Impact Information	A0705	Battery Problem		
					CT2750	Implantable cardiac electrical stimulators and associated devices	E0623	Unspecified Heart Problem	A0705	Battery Problem		
							F24	Insufficient Health Impact Information	A0705	Battery Problem		
					CT2885	Physiological monitoring/measuring system transmitters/receivers/transceivers	E0623	Unspecified Heart Problem	A0705	Battery Problem		
							F24	Insufficient Health Impact Information	A0705	Battery Problem		
17357	2025	All other reportable incidents	63188	Left-ventricular wireless pacing system	CT145	Cardiovascular devices	E0623	Unspecified Heart Problem	A0708	Power Problem		

				sensor/transmitter			F24	Insufficient Health Impact Information	A0708	Power Problem		
					CT1577	Wireless monitoring system transmitters	E0623	Unspecified Heart Problem	A0708	Power Problem		
							F24	Insufficient Health Impact Information	A0708	Power Problem		
					CT1684	Cardiac pacemakers and associated devices	E0623	Unspecified Heart Problem	A0708	Power Problem		
							F24	Insufficient Health Impact Information	A0708	Power Problem		
					CT2727	Active implant devices/ancillaries	E0623	Unspecified Heart Problem	A0708	Power Problem		
							F24	Insufficient Health Impact Information	A0708	Power Problem		
					CT2750	Implantable cardiac electrical stimulators and associated devices	E0623	Unspecified Heart Problem	A0708	Power Problem		
							F24	Insufficient Health Impact Information	A0708	Power Problem		
					CT2885	Physiological monitoring/measuring system transmitters/receivers/transceivers	E0623	Unspecified Heart Problem	A0708	Power Problem		
							F24	Insufficient Health Impact Information	A0708	Power Problem		
17358	2025	All other reportable incidents	37265	Dual-chamber implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A160601	Premature Elective Replacement Indicator		
							F1905	Device Revision or Replacement	A160601	Premature Elective Replacement Indicator		
17359	2025	All other reportable incidents	47265	Dual-chamber implantable pacemaker, rate-responsive	CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A070909	Oversensing		
									A072001	Electromagnetic Interference		
									A090801	Signal Artifact/Noise		
							F26	No Health Consequences or Impact	A070909	Oversensing		
									A072001	Electromagnetic Interference		
									A090801	Signal Artifact/Noise		
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A070909	Oversensing		
									A072001	Electromagnetic Interference		
									A090801	Signal Artifact/Noise		
							F26	No Health Consequences or Impact	A070909	Oversensing		
									A072001	Electromagnetic Interference		
									A090801	Signal Artifact/Noise		
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A070909	Oversensing		
									A072001	Electromagnetic Interference		
									A090801	Signal Artifact/Noise		
							F26	No Health Consequences or Impact	A070909	Oversensing		
									A072001	Electromagnetic Interference		
									A090801	Signal Artifact/Noise		
17360	2025	All other reportable incidents	47270	Cardiac resynchronization therapy implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A070101	Failure to Capture		
									A1301	Failure to Read Input Signal		
							F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A070101	Failure to Capture		
									A1301	Failure to Read Input Signal		
17361	2025	All other reportable incidents	37265	Dual-chamber implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A070909	Oversensing		
									A072001	Electromagnetic Interference		
							F26	No Health Consequences or Impact	A070909	Oversensing		
									A072001	Electromagnetic Interference		
17362	2025	All other reportable incidents	47265	Dual-chamber implantable pacemaker, rate-responsive	CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A070101	Failure to Capture		
									A0709	Device Sensing Problem		
							F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A070101	Failure to Capture		
									A0709	Device Sensing Problem		
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A070101	Failure to Capture		
									A0709	Device Sensing Problem		
							F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A070101	Failure to Capture		
									A0709	Device Sensing Problem		
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A070101	Failure to Capture		
									A0709	Device Sensing Problem		
							F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A070101	Failure to Capture		
									A0709	Device Sensing Problem		
17363	2025	All other reportable incidents	37265	Dual-chamber implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E060109	Tachycardia	A0709	Device Sensing Problem		
									A071301	Failure to Deliver Shock/Stimulation		
							F1001	Absence of Treatment	A0709	Device Sensing Problem		
									A071301	Failure to Deliver Shock/Stimulation		
							F1203	Life Threatening Illness or Injury	A0709	Device Sensing Problem		
									A071301	Failure to Deliver Shock/Stimulation		
							F2307	Reprogramming of Device	A0709	Device Sensing Problem		
									A071301	Failure to Deliver Shock/Stimulation		
17364	2025	All other reportable incidents	47265	Dual-chamber implantable pacemaker, rate-responsive	CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A0908	Incorrect, Inadequate or Imprecise Result or Readings	31/07/2025	31/07/2025
							F08	Hospitalization or Prolonged Hospitalization	A0908	Incorrect, Inadequate or Imprecise Result or Readings	31/07/2025	31/07/2025
							F1905	Device Revision or Replacement	A0908	Incorrect, Inadequate or Imprecise Result or Readings	31/07/2025	31/07/2025
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A0908	Incorrect, Inadequate or Imprecise Result or Readings	31/07/2025	31/07/2025
							F08	Hospitalization or Prolonged Hospitalization	A0908	Incorrect, Inadequate or Imprecise Result or Readings	31/07/2025	31/07/2025
							F1905	Device Revision or Replacement	A0908	Incorrect, Inadequate or Imprecise Result or Readings	31/07/2025	31/07/2025
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A0908	Incorrect, Inadequate or Imprecise Result or Readings	31/07/2025	31/07/2025
							F08	Hospitalization or Prolonged Hospitalization	A0908	Incorrect, Inadequate or Imprecise Result or Readings	31/07/2025	31/07/2025
							F1905	Device Revision or Replacement	A0908	Incorrect, Inadequate or Imprecise Result or Readings	31/07/2025	31/07/2025
17365	2025	All other reportable incidents	47267	Single-chamber implantable pacemaker, rate-responsive			E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	31/07/2025	31/07/2025
							F1901	Additional Surgery	A24	No Apparent Device Problem	31/07/2025	31/07/2025
17366	2025	All other reportable incidents	47265	Dual-chamber implantable pacemaker, rate-responsive			E2403	No Clinical Signs, Symptoms or Conditions	A070909	Oversensing		
									A1301	Failure to Read Input Signal		
							F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A070909	Oversensing		
									A1301	Failure to Read Input Signal		
17367	2025	All other reportable incidents	47265	Dual-chamber implantable pacemaker, rate-responsive	CT1685	Cardiac pacemakers	E060101	Asystole	A070909	Oversensing		
									A0712	Pacing Problem		
							F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A070909	Oversensing		
									A0712	Pacing Problem		
							F1001	Absence of Treatment	A070909	Oversensing		
									A0712	Pacing Problem		
					CT257	Implantable cardiac pacemakers	E060101	Asystole	A070909	Oversensing		

									A0712	Pacing Problem			
							F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A070909	Oversensing			
									A0712	Pacing Problem			
							F1001	Absence of Treatment	A070909	Oversensing			
									A0712	Pacing Problem			
					CT2866	Implantable cardiac electrical stimulators	E060101	Asystole	A070909	Oversensing			
									A0712	Pacing Problem			
							F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A070909	Oversensing			
									A0712	Pacing Problem			
							F1001	Absence of Treatment	A070909	Oversensing			
									A0712	Pacing Problem			
17368	2025	All other reportable incidents	47205	Cardiac pulse generator programmer	CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A13	Communication or Transmission Problem			
							F26	No Health Consequences or Impact	A13	Communication or Transmission Problem			
17369	2025	All other reportable incidents	47205	Cardiac pulse generator programmer	CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A13	Communication or Transmission Problem			
							F26	No Health Consequences or Impact	A13	Communication or Transmission Problem			
17370	2025	All other reportable incidents	47270	Cardiac resynchronization therapy implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A05	Mechanical Problem			
									A0712	Pacing Problem			
							F26	No Health Consequences or Impact	A05	Mechanical Problem			
									A0712	Pacing Problem			
17371	2025		37265	Dual-chamber implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E2401	Insufficient Information	A07	Electrical /Electronic Property Problem			
							F19	Surgical Intervention	A07	Electrical /Electronic Property Problem			
17372	2025	All other reportable incidents	47263	Cardiac resynchronization therapy implantable pacemaker	CT1685	Cardiac pacemakers	E0601	Arrhythmia	A24	No Apparent Device Problem	03/07/2025	03/07/2025	
							F23	Unexpected Medical Intervention	A24	No Apparent Device Problem	03/07/2025	03/07/2025	
					CT257	Implantable cardiac pacemakers	E0601	Arrhythmia	A24	No Apparent Device Problem	03/07/2025	03/07/2025	
							F23	Unexpected Medical Intervention	A24	No Apparent Device Problem	03/07/2025	03/07/2025	
					CT2866	Implantable cardiac electrical stimulators	E0601	Arrhythmia	A24	No Apparent Device Problem	03/07/2025	03/07/2025	
							F23	Unexpected Medical Intervention	A24	No Apparent Device Problem	03/07/2025	03/07/2025	
17373	2025	All other reportable incidents	47270	Cardiac resynchronization therapy implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E060104	Bradycardia	A0712	Pacing Problem			
							F11	Non-Serious Injury/ Illness/ Impairment	A0712	Pacing Problem			
17374	2025	All other reportable incidents	47265	Dual-chamber implantable pacemaker, rate-responsive	CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A0203	Defective Device	02/02/2025	02/02/2025	
							F24	Insufficient Health Impact Information	A0203	Defective Device	02/02/2025	02/02/2025	
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A0203	Defective Device	02/02/2025	02/02/2025	
							F24	Insufficient Health Impact Information	A0203	Defective Device	02/02/2025	02/02/2025	
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A0203	Defective Device	02/02/2025	02/02/2025	
							F24	Insufficient Health Impact Information	A0203	Defective Device	02/02/2025	02/02/2025	
17375	2025	All other reportable incidents	63188	Left-ventricular wireless pacing system sensor/transmitter	CT145	Cardiovascular devices	E0623	Unspecified Heart Problem	A0708	Power Problem			
							F24	Insufficient Health Impact Information	A0708	Power Problem			
					CT1577	Wireless monitoring system transmitters	E0623	Unspecified Heart Problem	A0708	Power Problem			
							F24	Insufficient Health Impact Information	A0708	Power Problem			
					CT1684	Cardiac pacemakers and associated devices	E0623	Unspecified Heart Problem	A0708	Power Problem			
							F24	Insufficient Health Impact Information	A0708	Power Problem			
					CT2727	Active implant devices/ancillaries	E0623	Unspecified Heart Problem	A0708	Power Problem			
							F24	Insufficient Health Impact Information	A0708	Power Problem			
					CT2750	Implantable cardiac electrical stimulators and associated devices	E0623	Unspecified Heart Problem	A0708	Power Problem			
							F24	Insufficient Health Impact Information	A0708	Power Problem			
					CT2885	Physiological monitoring/measuring system transmitters/receivers/transceivers	E0623	Unspecified Heart Problem	A0708	Power Problem			
							F24	Insufficient Health Impact Information	A0708	Power Problem			
17376	2025	Unanticipated serious deterioration in state of health	47265	Dual-chamber implantable pacemaker, rate-responsive	CT1685	Cardiac pacemakers	E2311	Discomfort	A23	Use of Device Problem			
							E2330	Pain	A23	Use of Device Problem			
							F12	Serious Injury/ Illness/ Impairment	A23	Use of Device Problem			
							F2307	Reprogramming of Device	A23	Use of Device Problem			
					CT257	Implantable cardiac pacemakers	E2311	Discomfort	A23	Use of Device Problem			
							E2330	Pain	A23	Use of Device Problem			
							F12	Serious Injury/ Illness/ Impairment	A23	Use of Device Problem			
							F2307	Reprogramming of Device	A23	Use of Device Problem			
					CT2866	Implantable cardiac electrical stimulators	E2311	Discomfort	A23	Use of Device Problem			
							E2330	Pain	A23	Use of Device Problem			
							F12	Serious Injury/ Illness/ Impairment	A23	Use of Device Problem			
							F2307	Reprogramming of Device	A23	Use of Device Problem			
17377	2025	All other reportable incidents	47270	Cardiac resynchronization therapy implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E210401	Shock from Patient Lead(s)	A070909	Oversensing			
									A071302	Inappropriate/Inadequate Shock/Stimulation			
									A1301	Failure to Read Input Signal			
							F08	Hospitalization or Prolonged Hospitalization	A070909	Oversensing			
									A071302	Inappropriate/Inadequate Shock/Stimulation			
									A1301	Failure to Read Input Signal			
							F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A070909	Oversensing			
									A071302	Inappropriate/Inadequate Shock/Stimulation			
									A1301	Failure to Read Input Signal			
17378	2025	All other reportable incidents	47270	Cardiac resynchronization therapy implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E210401	Shock from Patient Lead(s)	A0712	Pacing Problem			
									A071302	Inappropriate/Inadequate Shock/Stimulation			
							F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A0712	Pacing Problem			
									A071302	Inappropriate/Inadequate Shock/Stimulation			
17379	2025	All other reportable incidents	47270	Cardiac resynchronization therapy implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A070602	Delayed Charge Time			
									A070910	Undersensing			
							F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A070602	Delayed Charge Time			
									A070910	Undersensing			
17380	2025		60789	Intracardiac pacemaker	CT1685	Cardiac pacemakers	E0506	Hemorrhage/Blood Loss/Bleeding	A050401	Fluid/Blood Leak			
							F11	Non-Serious Injury/ Illness/ Impairment	A050401	Fluid/Blood Leak			
					CT257	Implantable cardiac pacemakers	E0506	Hemorrhage/Blood Loss/Bleeding	A050401	Fluid/Blood Leak			
							F11	Non-Serious Injury/ Illness/ Impairment	A050401	Fluid/Blood Leak			
					CT2818	Catheterization/Cannulation/Fluid path devices	E0506	Hemorrhage/Blood Loss/Bleeding	A050401	Fluid/Blood Leak			
							F11	Non-Serious Injury/ Illness/ Impairment	A050401	Fluid/Blood Leak			
					CT2866	Implantable cardiac electrical stimulators	E0506	Hemorrhage/Blood Loss/Bleeding	A050401	Fluid/Blood Leak			
							F11	Non-Serious Injury/ Illness/ Impairment	A050401	Fluid/Blood Leak			
17381	2025		60789	Intracardiac pacemaker	CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A21	Labeling, Instructions for Use or Training Problem			
									A210105	Missing Information			

							F24	Insufficient Health Impact Information	A21	Labeling, Instructions for Use or Training Problem		
									A210105	Missing Information		
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A21	Labeling, Instructions for Use or Training Problem		
									A210105	Missing Information		
							F24	Insufficient Health Impact Information	A21	Labeling, Instructions for Use or Training Problem		
									A210105	Missing Information		
					CT2818	Catheterization/Cannulation/Fluid path devices	E2403	No Clinical Signs, Symptoms or Conditions	A21	Labeling, Instructions for Use or Training Problem		
									A210105	Missing Information		
							F24	Insufficient Health Impact Information	A21	Labeling, Instructions for Use or Training Problem		
									A210105	Missing Information		
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A21	Labeling, Instructions for Use or Training Problem		
									A210105	Missing Information		
							F24	Insufficient Health Impact Information	A21	Labeling, Instructions for Use or Training Problem		
									A210105	Missing Information		
17382	2025	All other reportable incidents	47265	Dual-chamber implantable pacemaker, rate-responsive	CT1685	Cardiac pacemakers	E011903	Syncope/Fainting	A070101	Failure to Capture		
							F23	Unexpected Medical Intervention	A070101	Failure to Capture		
					CT257	Implantable cardiac pacemakers	E011903	Syncope/Fainting	A070101	Failure to Capture		
							F23	Unexpected Medical Intervention	A070101	Failure to Capture		
					CT2866	Implantable cardiac electrical stimulators	E011903	Syncope/Fainting	A070101	Failure to Capture		
							F23	Unexpected Medical Intervention	A070101	Failure to Capture		
17383	2025	Unanticipated serious deterioration in state of health	47265	Dual-chamber implantable pacemaker, rate-responsive	CT1685	Cardiac pacemakers	E060106	Heart Block	A0712	Pacing Problem		
							E0618	Non Specific EKG/ECG Changes	A0712	Pacing Problem		
							F12	Serious Injury/ Illness/ Impairment	A0712	Pacing Problem		
							F2307	Reprogramming of Device	A0712	Pacing Problem		
					CT257	Implantable cardiac pacemakers	E060106	Heart Block	A0712	Pacing Problem		
							E0618	Non Specific EKG/ECG Changes	A0712	Pacing Problem		
							F12	Serious Injury/ Illness/ Impairment	A0712	Pacing Problem		
							F2307	Reprogramming of Device	A0712	Pacing Problem		
					CT2866	Implantable cardiac electrical stimulators	E060106	Heart Block	A0712	Pacing Problem		
							E0618	Non Specific EKG/ECG Changes	A0712	Pacing Problem		
							F12	Serious Injury/ Illness/ Impairment	A0712	Pacing Problem		
							F2307	Reprogramming of Device	A0712	Pacing Problem		
17384	2025	All other reportable incidents	47205	Cardiac pulse generator programmer	CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A2202	Device Difficult to Program or Calibrate		
							F2601	Problem Identified Before Clinical Use/Exposure	A2202	Device Difficult to Program or Calibrate		
17385	2025	All other reportable incidents	37265	Dual-chamber implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A070102	High Capture Threshold		
									A070901	Decreased Sensitivity		
									A070909	Oversensing		
									A072001	Electromagnetic Interference		
							F26	No Health Consequences or Impact	A070102	High Capture Threshold		
									A070901	Decreased Sensitivity		
									A070909	Oversensing		
									A072001	Electromagnetic Interference		
17386	2025	All other reportable incidents	37265	Dual-chamber implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A090805	Incorrect Measurement	01/08/2025	01/08/2025
							F1901	Additional Surgery	A090805	Incorrect Measurement	01/08/2025	01/08/2025
17387	2025	All other reportable incidents	37265	Dual-chamber implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A072201	High Impedance		
							F26	No Health Consequences or Impact	A072201	High Impedance		
17388	2025	All other reportable incidents	35852	Single-chamber implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E210401	Shock from Patient Lead(s)	A070909	Oversensing		
									A071302	Inappropriate/Inadequate Shock/Stimulation		
							F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A070909	Oversensing		
									A071302	Inappropriate/Inadequate Shock/Stimulation		
							F2203	Imaging Required	A070909	Oversensing		
									A071302	Inappropriate/Inadequate Shock/Stimulation		
17389	2025	All other reportable incidents	35852	Single-chamber implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A070909	Oversensing		
									A0712	Pacing Problem		
									A1301	Failure to Read Input Signal		
							F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A070909	Oversensing		
									A0712	Pacing Problem		
									A1301	Failure to Read Input Signal		
17390	2025	All other reportable incidents	35852	Single-chamber implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E210401	Shock from Patient Lead(s)	A070909	Oversensing		
									A070910	Undersensing		
									A071302	Inappropriate/Inadequate Shock/Stimulation		
							F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A070909	Oversensing		
									A070910	Undersensing		
									A071302	Inappropriate/Inadequate Shock/Stimulation		
17391	2025	All other reportable incidents	37265	Dual-chamber implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A072201	High Impedance		
							F26	No Health Consequences or Impact	A072201	High Impedance		
17392	2025	All other reportable incidents	47263	Cardiac resynchronization therapy implantable pacemaker	CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A0203	Defective Device	01/03/2025	01/03/2025
							F24	Insufficient Health Impact Information	A0203	Defective Device	01/03/2025	01/03/2025
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A0203	Defective Device	01/03/2025	01/03/2025
							F24	Insufficient Health Impact Information	A0203	Defective Device	01/03/2025	01/03/2025
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A0203	Defective Device	01/03/2025	01/03/2025
							F24	Insufficient Health Impact Information	A0203	Defective Device	01/03/2025	01/03/2025
17393	2025	All other reportable incidents	35852	Single-chamber implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E210401	Shock from Patient Lead(s)	A070909	Oversensing		
									A071302	Inappropriate/Inadequate Shock/Stimulation		
							F08	Hospitalization or Prolonged Hospitalization	A070909	Oversensing		
									A071302	Inappropriate/Inadequate Shock/Stimulation		
							F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A070909	Oversensing		
									A071302	Inappropriate/Inadequate Shock/Stimulation		
							F23	Unexpected Medical Intervention	A070909	Oversensing		
									A071302	Inappropriate/Inadequate Shock/Stimulation		
17394	2025	All other reportable incidents	47265	Dual-chamber implantable pacemaker, rate-responsive	CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A0701	Capturing Problem		
							F26	No Health Consequences or Impact	A0701	Capturing Problem		

					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A0701	Capturing Problem		
							F26	No Health Consequences or Impact	A0701	Capturing Problem		
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A0701	Capturing Problem		
							F26	No Health Consequences or Impact	A0701	Capturing Problem		
17395	2025	All other reportable incidents	47265	Dual-chamber implantable pacemaker, rate-responsive			E060104	Bradycardia	A0712	Pacing Problem	15/08/2025	15/08/2025
									A0908	Incorrect, Inadequate or Imprecise Result or Readings	15/08/2025	15/08/2025
							F08	Hospitalization or Prolonged Hospitalization	A0712	Pacing Problem	15/08/2025	15/08/2025
									A0908	Incorrect, Inadequate or Imprecise Result or Readings	15/08/2025	15/08/2025
							F1905	Device Revision or Replacement	A0712	Pacing Problem	15/08/2025	15/08/2025
									A0908	Incorrect, Inadequate or Imprecise Result or Readings	15/08/2025	15/08/2025
17396	2025	All other reportable incidents	37265	Dual-chamber implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A072201	High Impedance		
							F26	No Health Consequences or Impact	A072201	High Impedance		
17397	2025	All other reportable incidents	35852	Single-chamber implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E210401	Shock from Patient Lead(s)	A0709	Device Sensing Problem		
									A070909	Oversensing		
									A071302	Inappropriate/Inadequate Shock/Stimulation		
							F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A0709	Device Sensing Problem		
									A070909	Oversensing		
									A071302	Inappropriate/Inadequate Shock/Stimulation		
17398	2025	All other reportable incidents	60956	Pacing/defibrillation lead tunneller	CT1684	Cardiac pacemakers and associated devices	E210401	Shock from Patient Lead(s)	A070909	Oversensing		
									A071302	Inappropriate/Inadequate Shock/Stimulation		
							F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A070909	Oversensing		
									A071302	Inappropriate/Inadequate Shock/Stimulation		
17399	2025	All other reportable incidents	47265	Dual-chamber implantable pacemaker, rate-responsive	CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A0908	Incorrect, Inadequate or Imprecise Result or Readings	19/06/2025	19/06/2025
							F08	Hospitalization or Prolonged Hospitalization	A0908	Incorrect, Inadequate or Imprecise Result or Readings	19/06/2025	19/06/2025
							F1905	Device Revision or Replacement	A0908	Incorrect, Inadequate or Imprecise Result or Readings	19/06/2025	19/06/2025
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A0908	Incorrect, Inadequate or Imprecise Result or Readings	19/06/2025	19/06/2025
							F08	Hospitalization or Prolonged Hospitalization	A0908	Incorrect, Inadequate or Imprecise Result or Readings	19/06/2025	19/06/2025
							F1905	Device Revision or Replacement	A0908	Incorrect, Inadequate or Imprecise Result or Readings	19/06/2025	19/06/2025
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A0908	Incorrect, Inadequate or Imprecise Result or Readings	19/06/2025	19/06/2025
							F08	Hospitalization or Prolonged Hospitalization	A0908	Incorrect, Inadequate or Imprecise Result or Readings	19/06/2025	19/06/2025
							F1905	Device Revision or Replacement	A0908	Incorrect, Inadequate or Imprecise Result or Readings	19/06/2025	19/06/2025
17400	2025	All other reportable incidents	35852	Single-chamber implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E0601	Arrhythmia	A160503	Inappropriate or Unexpected Reset		
							F1901	Additional Surgery	A160503	Inappropriate or Unexpected Reset		
17401	2025	All other reportable incidents	17882	Physiologic-monitoring defibrillation system	CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A0801	Failure to Calibrate		
							F27	Problem Identified During Non-Clinical Procedure	A0801	Failure to Calibrate		
17402	2025	All other reportable incidents	35852	Single-chamber implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E1906	Unspecified Infection	A070504	Premature Discharge of Battery		
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery		
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery		
17403	2025	All other reportable incidents	35852	Single-chamber implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A0709	Device Sensing Problem		
									A070909	Oversensing		
									A090801	Signal Artifact/Noise		
							F08	Hospitalization or Prolonged Hospitalization	A0709	Device Sensing Problem		
									A070909	Oversensing		
									A090801	Signal Artifact/Noise		
17404	2025	All other reportable incidents	47270	Cardiac resynchronization therapy implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A0709	Device Sensing Problem		
									A070909	Oversensing		
									A090801	Signal Artifact/Noise		
							F08	Hospitalization or Prolonged Hospitalization	A0709	Device Sensing Problem		
									A070909	Oversensing		
									A090801	Signal Artifact/Noise		
17405	2025	All other reportable incidents	37265	Dual-chamber implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E210401	Shock from Patient Lead(s)	A071302	Inappropriate/Inadequate Shock/Stimulation		
							F08	Hospitalization or Prolonged Hospitalization	A071302	Inappropriate/Inadequate Shock/Stimulation		
							F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A071302	Inappropriate/Inadequate Shock/Stimulation		
17406	2025	All other reportable incidents	37265	Dual-chamber implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A072201	High Impedance		
									A090801	Signal Artifact/Noise		
							F2203	Imaging Required	A072201	High Impedance		
									A090801	Signal Artifact/Noise		
							F26	No Health Consequences or Impact	A072201	High Impedance		
									A090801	Signal Artifact/Noise		
17407	2025	All other reportable incidents	47270	Cardiac resynchronization therapy implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A12	Connection Problem		
							F2203	Imaging Required	A12	Connection Problem		
17408	2025		60956	Pacing/defibrillation lead tunneller	CT1684	Cardiac pacemakers and associated devices	E2401	Insufficient Information	A0705	Battery Problem		
							F1901	Additional Surgery	A0705	Battery Problem		
17409	2025		47265	Dual-chamber implantable pacemaker, rate-responsive	CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A04	Material Integrity Problem		
									A090801	Signal Artifact/Noise		
							F26	No Health Consequences or Impact	A04	Material Integrity Problem		
									A090801	Signal Artifact/Noise		
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A04	Material Integrity Problem		
									A090801	Signal Artifact/Noise		
							F26	No Health Consequences or Impact	A04	Material Integrity Problem		
									A090801	Signal Artifact/Noise		
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A04	Material Integrity Problem		
									A090801	Signal Artifact/Noise		
							F26	No Health Consequences or Impact	A04	Material Integrity Problem		
									A090801	Signal Artifact/Noise		
17410	2025	All other reportable incidents	47265	Dual-chamber implantable pacemaker, rate-responsive			E2403	No Clinical Signs, Symptoms or Conditions	A0908	Incorrect, Inadequate or Imprecise Result or Readings	09/01/2026	09/01/2026
							F08	Hospitalization or Prolonged Hospitalization	A0908	Incorrect, Inadequate or Imprecise Result or Readings	09/01/2026	09/01/2026

							F1905	Device Revision or Replacement	A0908	Incorrect, Inadequate or Imprecise Result or Readings	09/01/2026	09/01/2026
17411	2025	All other reportable incidents	35852	Single-chamber implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem	20/09/2024	20/09/2024
									A070504	Premature Discharge of Battery	20/09/2024	20/09/2024
									A0908	Incorrect, Inadequate or Imprecise Result or Readings	20/09/2024	20/09/2024
							F08	Hospitalization or Prolonged Hospitalization	A0705	Battery Problem	20/09/2024	20/09/2024
									A070504	Premature Discharge of Battery	20/09/2024	20/09/2024
									A0908	Incorrect, Inadequate or Imprecise Result or Readings	20/09/2024	20/09/2024
							F1905	Device Revision or Replacement	A0705	Battery Problem	20/09/2024	20/09/2024
									A070504	Premature Discharge of Battery	20/09/2024	20/09/2024
									A0908	Incorrect, Inadequate or Imprecise Result or Readings	20/09/2024	20/09/2024
17412	2025	All other reportable incidents	47270	Cardiac resynchronization therapy implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A070909	Oversensing		
							F26	No Health Consequences or Impact	A070909	Oversensing		
17413	2025	All other reportable incidents	47265	Dual-chamber implantable pacemaker, rate-responsive	CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
17414	2025	All other reportable incidents	35852	Single-chamber implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A072201	High Impedance		
							F26	No Health Consequences or Impact	A072201	High Impedance		
17415	2025	All other reportable incidents	47270	Cardiac resynchronization therapy implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E210401	Shock from Patient Lead(s)	A070909	Oversensing		
									A071302	Inappropriate/Inadequate Shock/Stimulation		
									A090801	Signal Artifact/Noise		
							F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A070909	Oversensing		
									A071302	Inappropriate/Inadequate Shock/Stimulation		
									A090801	Signal Artifact/Noise		
17416	2025	All other reportable incidents	37265	Dual-chamber implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A070909	Oversensing		
							F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A070909	Oversensing		
17417	2025	All other reportable incidents	36106	Implantable pacemaker analyser, intraoperative	CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A11	Computer Software Problem		
									A13	Communication or Transmission Problem		
							F26	No Health Consequences or Impact	A11	Computer Software Problem		
									A13	Communication or Transmission Problem		
17418	2025		47265	Dual-chamber implantable pacemaker, rate-responsive	CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A071206	Pacing Asynchronously		
							F26	No Health Consequences or Impact	A071206	Pacing Asynchronously		
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A071206	Pacing Asynchronously		
							F26	No Health Consequences or Impact	A071206	Pacing Asynchronously		
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A071206	Pacing Asynchronously		
							F26	No Health Consequences or Impact	A071206	Pacing Asynchronously		
17419	2025		60956	Pacing/defibrillation lead tunneller	CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A071102	Failure to Interrogate		
							F1903	Device Explantation	A071102	Failure to Interrogate		
17420	2025	All other reportable incidents	47270	Cardiac resynchronization therapy implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A1304	Telemetry Discrepancy		
							F26	No Health Consequences or Impact	A1304	Telemetry Discrepancy		
17421	2025	All other reportable incidents	36106	Implantable pacemaker analyser, intraoperative	CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A13	Communication or Transmission Problem		
							F26	No Health Consequences or Impact	A13	Communication or Transmission Problem		
17422	2025	All other reportable incidents	36106	Implantable pacemaker analyser, intraoperative	CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A0709	Device Sensing Problem		
									A13	Communication or Transmission Problem		
									A2204	Inadequate User Interface		
							F26	No Health Consequences or Impact	A0709	Device Sensing Problem		
									A13	Communication or Transmission Problem		
									A2204	Inadequate User Interface		
17423	2025	All other reportable incidents	36106	Implantable pacemaker analyser, intraoperative	CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A0709	Device Sensing Problem		
									A13	Communication or Transmission Problem		
									A2204	Inadequate User Interface		
							F26	No Health Consequences or Impact	A0709	Device Sensing Problem		
									A13	Communication or Transmission Problem		
									A2204	Inadequate User Interface		
17424	2025	All other reportable incidents	36106	Implantable pacemaker analyser, intraoperative	CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A0709	Device Sensing Problem		
									A13	Communication or Transmission Problem		
									A2204	Inadequate User Interface		
							F26	No Health Consequences or Impact	A0709	Device Sensing Problem		
									A13	Communication or Transmission Problem		
									A2204	Inadequate User Interface		
17425	2025	All other reportable incidents	35852	Single-chamber implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A1605	Reset Problem		
									A160503	Inappropriate or Unexpected Reset		
							F26	No Health Consequences or Impact	A1605	Reset Problem		
									A160503	Inappropriate or Unexpected Reset		
17426	2025	All other reportable incidents	35852	Single-chamber implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A072202	Low Impedance		
									A090801	Signal Artifact/Noise		
							F26	No Health Consequences or Impact	A072202	Low Impedance		
									A090801	Signal Artifact/Noise		
17427	2025	All other reportable incidents	47270	Cardiac resynchronization therapy implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
17428	2025	All other reportable incidents	37265	Dual-chamber implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A070909	Oversensing		
							F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A070909	Oversensing		
17429	2025	All other reportable incidents	35852	Single-chamber implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A072201	High Impedance		
							F26	No Health Consequences or Impact	A072201	High Impedance		
17430	2025	All other reportable incidents	47265	Dual-chamber implantable pacemaker, rate-responsive			E2403	No Clinical Signs, Symptoms or Conditions	A0908	Incorrect, Inadequate or Imprecise Result or Readings	18/08/2025	18/08/2025
							F08	Hospitalization or Prolonged Hospitalization	A0908	Incorrect, Inadequate or Imprecise Result or Readings	18/08/2025	18/08/2025
							F1905	Device Revision or Replacement	A0908	Incorrect, Inadequate or Imprecise Result or Readings	18/08/2025	18/08/2025
17431	2025	All other reportable incidents	47270	Cardiac resynchronization therapy implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A072201	High Impedance		
									A1301	Failure to Read Input Signal		
							F1904	Device Repositioning	A072201	High Impedance		

									A1301	Failure to Read Input Signal		
							F1908	Prolonged Surgery	A072201	High Impedance		
									A1301	Failure to Read Input Signal		
							F23	Unexpected Medical Intervention	A072201	High Impedance		
									A1301	Failure to Read Input Signal		
17432	2025	All other reportable incidents	47270	Cardiac resynchronization therapy implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A072201	High Impedance		
									A1301	Failure to Read Input Signal		
							F1904	Device Repositioning	A072201	High Impedance		
									A1301	Failure to Read Input Signal		
							F1908	Prolonged Surgery	A072201	High Impedance		
									A1301	Failure to Read Input Signal		
							F23	Unexpected Medical Intervention	A072201	High Impedance		
									A1301	Failure to Read Input Signal		
17433	2025	All other reportable incidents	35852	Single-chamber implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A070912	Incorrect Interpretation of Signal		
									A071302	Inappropriate/Inadequate Shock/Stimulation		
							F12	Serious Injury/ Illness/ Impairment	A070912	Incorrect Interpretation of Signal		
									A071302	Inappropriate/Inadequate Shock/Stimulation		
17434	2025	All other reportable incidents	35852	Single-chamber implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
									A0908	Incorrect, Inadequate or Imprecise Result or Readings		
							F26	No Health Consequences or Impact	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
									A0908	Incorrect, Inadequate or Imprecise Result or Readings		
17435	2025	All other reportable incidents	36106	Implantable pacemaker analyser, intraoperative	CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A071102	Failure to Interrogate		
									A1104	Computer Operating System Problem		
							F26	No Health Consequences or Impact	A071102	Failure to Interrogate		
									A1104	Computer Operating System Problem		
17436	2025	All other reportable incidents	36106	Implantable pacemaker analyser, intraoperative	CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A0709	Device Sensing Problem		
									A071102	Failure to Interrogate		
									A1104	Computer Operating System Problem		
							F26	No Health Consequences or Impact	A0709	Device Sensing Problem		
									A071102	Failure to Interrogate		
									A1104	Computer Operating System Problem		
17437	2025	All other reportable incidents	36106	Implantable pacemaker analyser, intraoperative	CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A13	Communication or Transmission Problem		
							F26	No Health Consequences or Impact	A13	Communication or Transmission Problem		
17438	2025	All other reportable incidents	36106	Implantable pacemaker analyser, intraoperative	CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A0709	Device Sensing Problem		
									A09	Output Problem		
									A11	Computer Software Problem		
							F26	No Health Consequences or Impact	A0709	Device Sensing Problem		
									A09	Output Problem		
									A11	Computer Software Problem		
17439	2025	All other reportable incidents	47263	Cardiac resynchronization therapy implantable pacemaker	CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A0908	Incorrect, Inadequate or Imprecise Result or Readings	27/08/2025	27/08/2025
							F08	Hospitalization or Prolonged Hospitalization	A0908	Incorrect, Inadequate or Imprecise Result or Readings	27/08/2025	27/08/2025
							F1905	Device Revision or Replacement	A0908	Incorrect, Inadequate or Imprecise Result or Readings	27/08/2025	27/08/2025
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A0908	Incorrect, Inadequate or Imprecise Result or Readings	27/08/2025	27/08/2025
							F08	Hospitalization or Prolonged Hospitalization	A0908	Incorrect, Inadequate or Imprecise Result or Readings	27/08/2025	27/08/2025
							F1905	Device Revision or Replacement	A0908	Incorrect, Inadequate or Imprecise Result or Readings	27/08/2025	27/08/2025
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A0908	Incorrect, Inadequate or Imprecise Result or Readings	27/08/2025	27/08/2025
							F08	Hospitalization or Prolonged Hospitalization	A0908	Incorrect, Inadequate or Imprecise Result or Readings	27/08/2025	27/08/2025
							F1905	Device Revision or Replacement	A0908	Incorrect, Inadequate or Imprecise Result or Readings	27/08/2025	27/08/2025
17440	2025	All other reportable incidents	47265	Dual-chamber implantable pacemaker, rate-responsive			E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	12/08/2025	12/08/2025
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	12/08/2025	12/08/2025
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	12/08/2025	12/08/2025
17441	2025	All other reportable incidents	47267	Single-chamber implantable pacemaker, rate-responsive	CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A071102	Failure to Interrogate	28/08/2025	28/08/2025
							F1901	Additional Surgery	A071102	Failure to Interrogate	28/08/2025	28/08/2025
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A071102	Failure to Interrogate	28/08/2025	28/08/2025
							F1901	Additional Surgery	A071102	Failure to Interrogate	28/08/2025	28/08/2025
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A071102	Failure to Interrogate	28/08/2025	28/08/2025
							F1901	Additional Surgery	A071102	Failure to Interrogate	28/08/2025	28/08/2025
17442	2025	All other reportable incidents	35852	Single-chamber implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A070909	Oversensing		
									A070910	Undersensing		
							F26	No Health Consequences or Impact	A070909	Oversensing		
									A070910	Undersensing		
17443	2025	All other reportable incidents	47265	Dual-chamber implantable pacemaker, rate-responsive	CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A071102	Failure to Interrogate		
							F1901	Additional Surgery	A071102	Failure to Interrogate		
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A071102	Failure to Interrogate		
							F1901	Additional Surgery	A071102	Failure to Interrogate		
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A071102	Failure to Interrogate		
							F1901	Additional Surgery	A071102	Failure to Interrogate		
17444	2025		57968	Cardiovascular device introducer, steerable	CT1684	Cardiac pacemakers and associated devices	E0601	Arrhythmia	A040607	Material Too Rigid or Stiff		
							E0604	Cardiac Perforation	A040607	Material Too Rigid or Stiff		
							E0605	Cardiac Tamponade	A040607	Material Too Rigid or Stiff		
							E0619	Pericardial Effusion	A040607	Material Too Rigid or Stiff		
							E2402	Appropriate Clinical Signs, Symptoms and Conditions Term/Code not Available	A040607	Material Too Rigid or Stiff		
							F02	Death	A040607	Material Too Rigid or Stiff		
17445	2025	All other reportable incidents	36106	Implantable pacemaker analyser, intraoperative	CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A13	Communication or Transmission Problem		
							F26	No Health Consequences or Impact	A13	Communication or Transmission Problem		
17446	2025	All other reportable incidents	60956	Pacing/defibrillation lead tunneller	CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070909	Oversensing		
							F26	No Health Consequences or Impact	A070909	Oversensing		

17447	2025	All other reportable incidents	47265	Dual-chamber implantable pacemaker, rate-responsive			E2403	No Clinical Signs, Symptoms or Conditions	A0908	Incorrect, Inadequate or Imprecise Result or Readings	15/09/2025	15/09/2025
							F08	Hospitalization or Prolonged Hospitalization	A0908	Incorrect, Inadequate or Imprecise Result or Readings	15/09/2025	15/09/2025
							F1905	Device Revision or Replacement	A0908	Incorrect, Inadequate or Imprecise Result or Readings	15/09/2025	15/09/2025
17448	2025		47265	Dual-chamber implantable pacemaker, rate-responsive	CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	08/09/2025	08/09/2025
							F1901	Additional Surgery	A070504	Premature Discharge of Battery	08/09/2025	08/09/2025
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	08/09/2025	08/09/2025
							F1901	Additional Surgery	A070504	Premature Discharge of Battery	08/09/2025	08/09/2025
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	08/09/2025	08/09/2025
							F1901	Additional Surgery	A070504	Premature Discharge of Battery	08/09/2025	08/09/2025
17449	2025		47265	Dual-chamber implantable pacemaker, rate-responsive	CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem		
							F1901	Additional Surgery	A24	No Apparent Device Problem		
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem		
							F1901	Additional Surgery	A24	No Apparent Device Problem		
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem		
							F1901	Additional Surgery	A24	No Apparent Device Problem		
17450	2025		60789	Intracardiac pacemaker	CT1685	Cardiac pacemakers	E012207	Twitching	A0712	Pacing Problem		
									A0723	Unexpected Mode Switch		
							F01	Change in Therapeutic Response	A0712	Pacing Problem		
									A0723	Unexpected Mode Switch		
					CT257	Implantable cardiac pacemakers	E012207	Twitching	A0712	Pacing Problem		
									A0723	Unexpected Mode Switch		
							F01	Change in Therapeutic Response	A0712	Pacing Problem		
									A0723	Unexpected Mode Switch		
					CT2818	Catheterization/Cannulation/Fluid path devices	E012207	Twitching	A0712	Pacing Problem		
									A0723	Unexpected Mode Switch		
							F01	Change in Therapeutic Response	A0712	Pacing Problem		
									A0723	Unexpected Mode Switch		
					CT2866	Implantable cardiac electrical stimulators	E012207	Twitching	A0712	Pacing Problem		
									A0723	Unexpected Mode Switch		
							F01	Change in Therapeutic Response	A0712	Pacing Problem		
									A0723	Unexpected Mode Switch		
17451	2025		61499	Vascular catheter introduction kit, steerable, single-use	CT1684	Cardiac pacemakers and associated devices	E233001	Chest Pain	A0413	Material Separation		
							F1901	Additional Surgery	A0413	Material Separation		
							F1903	Device Explantation	A0413	Material Separation		
17452	2025	All other reportable incidents	47270	Cardiac resynchronization therapy implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A0709	Device Sensing Problem		
									A070909	Oversensing		
									A090801	Signal Artifact/Noise		
							F26	No Health Consequences or Impact	A0709	Device Sensing Problem		
									A070909	Oversensing		
									A090801	Signal Artifact/Noise		
17453	2025	All other reportable incidents	47270	Cardiac resynchronization therapy implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A071202	Inaccurate Synchronization		
									A0909	Unexpected Therapeutic Results		
							F15	Recognized Device or Procedural Complication	A071202	Inaccurate Synchronization		
									A0909	Unexpected Therapeutic Results		
17454	2025	All other reportable incidents	47265	Dual-chamber implantable pacemaker, rate-responsive	CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
									A0908	Incorrect, Inadequate or Imprecise Result or Readings		
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery		
									A0908	Incorrect, Inadequate or Imprecise Result or Readings		
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
									A0908	Incorrect, Inadequate or Imprecise Result or Readings		
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery		
									A0908	Incorrect, Inadequate or Imprecise Result or Readings		
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
									A0908	Incorrect, Inadequate or Imprecise Result or Readings		
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery		
									A0908	Incorrect, Inadequate or Imprecise Result or Readings		
17455	2025		17882	Physiologic-monitoring defibrillation system	CT1684	Cardiac pacemakers and associated devices	E1701	Abrasion	A0101	Patient-Device Incompatibility		
									A010101	Biocompatibility		
							F11	Non-Serious Injury/ Illness/ Impairment	A0101	Patient-Device Incompatibility		
									A010101	Biocompatibility		
17456	2025		47265	Dual-chamber implantable pacemaker, rate-responsive	CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A071208	Pacing Intermittently	22/08/2025	22/08/2025
							F1901	Additional Surgery	A071208	Pacing Intermittently	22/08/2025	22/08/2025
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A071208	Pacing Intermittently	22/08/2025	22/08/2025
							F1901	Additional Surgery	A071208	Pacing Intermittently	22/08/2025	22/08/2025
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A071208	Pacing Intermittently	22/08/2025	22/08/2025
							F1901	Additional Surgery	A071208	Pacing Intermittently	22/08/2025	22/08/2025
17457	2025	All other reportable incidents	37265	Dual-chamber implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A072201	High Impedance		
							F26	No Health Consequences or Impact	A072201	High Impedance		
17458	2025	All other reportable incidents	47263	Cardiac resynchronization therapy implantable pacemaker	CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A0908	Incorrect, Inadequate or Imprecise Result or Readings	26/08/2025	26/08/2025
							F08	Hospitalization or Prolonged Hospitalization	A0908	Incorrect, Inadequate or Imprecise Result or Readings	26/08/2025	26/08/2025
							F1905	Device Revision or Replacement	A0908	Incorrect, Inadequate or Imprecise Result or Readings	26/08/2025	26/08/2025
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A0908	Incorrect, Inadequate or Imprecise Result or Readings	26/08/2025	26/08/2025
							F08	Hospitalization or Prolonged Hospitalization	A0908	Incorrect, Inadequate or Imprecise Result or Readings	26/08/2025	26/08/2025
							F1905	Device Revision or Replacement	A0908	Incorrect, Inadequate or Imprecise Result or Readings	26/08/2025	26/08/2025
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A0908	Incorrect, Inadequate or Imprecise Result or Readings	26/08/2025	26/08/2025

							F08	Hospitalization or Prolonged Hospitalization	A0908	Incorrect, Inadequate or Imprecise Result or Readings	26/08/2025	26/08/2025
							F1905	Device Revision or Replacement	A0908	Incorrect, Inadequate or Imprecise Result or Readings	26/08/2025	26/08/2025
17459	2025	All other reportable incidents	47265	Dual-chamber implantable pacemaker, rate-responsive	CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
17460	2025	All other reportable incidents	17882	Physiologic-monitoring defibrillation system	CT1684	Cardiac pacemakers and associated devices	E0623	Unspecified Heart Problem	A0908	Incorrect, Inadequate or Imprecise Result or Readings		
							F05	Delay to Treatment/ Therapy	A0908	Incorrect, Inadequate or Imprecise Result or Readings		
17461	2025	All other reportable incidents	63188	Left-ventricular wireless pacing system sensor/transmitter	CT145	Cardiovascular devices	E0623	Unspecified Heart Problem	A25	No Apparent Adverse Event	13/06/2025	13/06/2025
							F24	Insufficient Health Impact Information	A25	No Apparent Adverse Event	13/06/2025	13/06/2025
					CT1577	Wireless monitoring system transmitters	E0623	Unspecified Heart Problem	A25	No Apparent Adverse Event	13/06/2025	13/06/2025
							F24	Insufficient Health Impact Information	A25	No Apparent Adverse Event	13/06/2025	13/06/2025
					CT1684	Cardiac pacemakers and associated devices	E0623	Unspecified Heart Problem	A25	No Apparent Adverse Event	13/06/2025	13/06/2025
							F24	Insufficient Health Impact Information	A25	No Apparent Adverse Event	13/06/2025	13/06/2025
					CT2727	Active implant devices/ancillaries	E0623	Unspecified Heart Problem	A25	No Apparent Adverse Event	13/06/2025	13/06/2025
							F24	Insufficient Health Impact Information	A25	No Apparent Adverse Event	13/06/2025	13/06/2025
					CT2750	Implantable cardiac electrical stimulators and associated devices	E0623	Unspecified Heart Problem	A25	No Apparent Adverse Event	13/06/2025	13/06/2025
							F24	Insufficient Health Impact Information	A25	No Apparent Adverse Event	13/06/2025	13/06/2025
					CT2885	Physiological monitoring/measuring system transmitters/receivers/transceivers	E0623	Unspecified Heart Problem	A25	No Apparent Adverse Event	13/06/2025	13/06/2025
							F24	Insufficient Health Impact Information	A25	No Apparent Adverse Event	13/06/2025	13/06/2025
17462	2025	All other reportable incidents	63188	Left-ventricular wireless pacing system sensor/transmitter	CT145	Cardiovascular devices	E0623	Unspecified Heart Problem	A0708	Power Problem		
							F24	Insufficient Health Impact Information	A0708	Power Problem		
					CT1577	Wireless monitoring system transmitters	E0623	Unspecified Heart Problem	A0708	Power Problem		
							F24	Insufficient Health Impact Information	A0708	Power Problem		
					CT1684	Cardiac pacemakers and associated devices	E0623	Unspecified Heart Problem	A0708	Power Problem		
							F24	Insufficient Health Impact Information	A0708	Power Problem		
					CT2727	Active implant devices/ancillaries	E0623	Unspecified Heart Problem	A0708	Power Problem		
							F24	Insufficient Health Impact Information	A0708	Power Problem		
					CT2750	Implantable cardiac electrical stimulators and associated devices	E0623	Unspecified Heart Problem	A0708	Power Problem		
							F24	Insufficient Health Impact Information	A0708	Power Problem		
					CT2885	Physiological monitoring/measuring system transmitters/receivers/transceivers	E0623	Unspecified Heart Problem	A0708	Power Problem		
							F24	Insufficient Health Impact Information	A0708	Power Problem		
17463	2025	All other reportable incidents	17882	Physiologic-monitoring defibrillation system	CT1684	Cardiac pacemakers and associated devices	E0623	Unspecified Heart Problem	A2303	Improper or Incorrect Procedure or Method		
							F02	Death	A2303	Improper or Incorrect Procedure or Method		
17464	2025	All other reportable incidents	60789	Intracardiac pacemaker	CT1685	Cardiac pacemakers	E0618	Non Specific EKG/ECG Changes	A070102	High Capture Threshold		
									A0712	Pacing Problem		
									A23	Use of Device Problem		
							E2403	No Clinical Signs, Symptoms or Conditions	A070102	High Capture Threshold		
									A0712	Pacing Problem		
									A23	Use of Device Problem		
							F12	Serious Injury/ Illness/ Impairment	A070102	High Capture Threshold		
									A0712	Pacing Problem		
									A23	Use of Device Problem		
							F2307	Reprogramming of Device	A070102	High Capture Threshold		
									A0712	Pacing Problem		
									A23	Use of Device Problem		
							F26	No Health Consequences or Impact	A070102	High Capture Threshold		
									A0712	Pacing Problem		
									A23	Use of Device Problem		
					CT257	Implantable cardiac pacemakers	E0618	Non Specific EKG/ECG Changes	A070102	High Capture Threshold		
									A0712	Pacing Problem		
									A23	Use of Device Problem		
							E2403	No Clinical Signs, Symptoms or Conditions	A070102	High Capture Threshold		
									A0712	Pacing Problem		
									A23	Use of Device Problem		
							F12	Serious Injury/ Illness/ Impairment	A070102	High Capture Threshold		
									A0712	Pacing Problem		
									A23	Use of Device Problem		
							F2307	Reprogramming of Device	A070102	High Capture Threshold		
									A0712	Pacing Problem		
									A23	Use of Device Problem		
							F26	No Health Consequences or Impact	A070102	High Capture Threshold		
									A0712	Pacing Problem		
									A23	Use of Device Problem		
					CT2818	Catheterization/Cannulation/Fluid path devices	E0618	Non Specific EKG/ECG Changes	A070102	High Capture Threshold		
									A0712	Pacing Problem		
									A23	Use of Device Problem		
							E2403	No Clinical Signs, Symptoms or Conditions	A070102	High Capture Threshold		
									A0712	Pacing Problem		
									A23	Use of Device Problem		
							F12	Serious Injury/ Illness/ Impairment	A070102	High Capture Threshold		
									A0712	Pacing Problem		
									A23	Use of Device Problem		
							F2307	Reprogramming of Device	A070102	High Capture Threshold		
									A0712	Pacing Problem		
									A23	Use of Device Problem		
							F26	No Health Consequences or Impact	A070102	High Capture Threshold		

									A0712	Pacing Problem		
									A23	Use of Device Problem		
					CT2866	Implantable cardiac electrical stimulators	E0618	Non Specific EKG/ECG Changes	A070102	High Capture Threshold		
									A0712	Pacing Problem		
									A23	Use of Device Problem		
							E2403	No Clinical Signs, Symptoms or Conditions	A070102	High Capture Threshold		
									A0712	Pacing Problem		
									A23	Use of Device Problem		
							F12	Serious Injury/ Illness/ Impairment	A070102	High Capture Threshold		
									A0712	Pacing Problem		
									A23	Use of Device Problem		
							F2307	Reprogramming of Device	A070102	High Capture Threshold		
									A0712	Pacing Problem		
									A23	Use of Device Problem		
							F26	No Health Consequences or Impact	A070102	High Capture Threshold		
									A0712	Pacing Problem		
									A23	Use of Device Problem		
17465	2025	All other reportable incidents	60789	Intracardiac pacemaker	CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A070102	High Capture Threshold	29/08/2025	29/08/2025
									A070909	Oversensing	29/08/2025	29/08/2025
							F26	No Health Consequences or Impact	A070102	High Capture Threshold	29/08/2025	29/08/2025
									A070909	Oversensing	29/08/2025	29/08/2025
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A070102	High Capture Threshold	29/08/2025	29/08/2025
									A070909	Oversensing	29/08/2025	29/08/2025
							F26	No Health Consequences or Impact	A070102	High Capture Threshold	29/08/2025	29/08/2025
									A070909	Oversensing	29/08/2025	29/08/2025
					CT2818	Catheterization/Cannulation/Fluid path devices	E2403	No Clinical Signs, Symptoms or Conditions	A070102	High Capture Threshold	29/08/2025	29/08/2025
									A070909	Oversensing	29/08/2025	29/08/2025
							F26	No Health Consequences or Impact	A070102	High Capture Threshold	29/08/2025	29/08/2025
									A070909	Oversensing	29/08/2025	29/08/2025
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A070102	High Capture Threshold	29/08/2025	29/08/2025
									A070909	Oversensing	29/08/2025	29/08/2025
							F26	No Health Consequences or Impact	A070102	High Capture Threshold	29/08/2025	29/08/2025
									A070909	Oversensing	29/08/2025	29/08/2025
17466	2025	All other reportable incidents	36106	Implantable pacemaker analyser, intraoperative	CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A11	Computer Software Problem		
									A1104	Computer Operating System Problem		
							F26	No Health Consequences or Impact	A11	Computer Software Problem		
									A1104	Computer Operating System Problem		
17467	2025	All other reportable incidents	35852	Single-chamber implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
									A0908	Incorrect, Inadequate or Imprecise Result or Readings		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
									A0908	Incorrect, Inadequate or Imprecise Result or Readings		
17468	2025	Unanticipated serious deterioration in state of health	47265	Dual-chamber implantable pacemaker, rate-responsive	CT1685	Cardiac pacemakers	E011903	Syncope/Fainting	A0712	Pacing Problem		
							E060104	Bradycardia	A0712	Pacing Problem		
							E0618	Non Specific EKG/ECG Changes	A0712	Pacing Problem		
							F08	Hospitalization or Prolonged Hospitalization	A0712	Pacing Problem		
							F1203	Life Threatening Illness or Injury	A0712	Pacing Problem		
							F2307	Reprogramming of Device	A0712	Pacing Problem		
					CT257	Implantable cardiac pacemakers	E011903	Syncope/Fainting	A0712	Pacing Problem		
							E060104	Bradycardia	A0712	Pacing Problem		
							E0618	Non Specific EKG/ECG Changes	A0712	Pacing Problem		
							F08	Hospitalization or Prolonged Hospitalization	A0712	Pacing Problem		
							F1203	Life Threatening Illness or Injury	A0712	Pacing Problem		
							F2307	Reprogramming of Device	A0712	Pacing Problem		
					CT2866	Implantable cardiac electrical stimulators	E011903	Syncope/Fainting	A0712	Pacing Problem		
							E060104	Bradycardia	A0712	Pacing Problem		
							E0618	Non Specific EKG/ECG Changes	A0712	Pacing Problem		
							F08	Hospitalization or Prolonged Hospitalization	A0712	Pacing Problem		
							F1203	Life Threatening Illness or Injury	A0712	Pacing Problem		
							F2307	Reprogramming of Device	A0712	Pacing Problem		
17469	2025	All other reportable incidents	60789	Intracardiac pacemaker			E2403	No Clinical Signs, Symptoms or Conditions	A070102	High Capture Threshold	08/08/2025	08/08/2025
									A1502	Positioning Problem	08/08/2025	08/08/2025
									A23	Use of Device Problem	08/08/2025	08/08/2025
							F26	No Health Consequences or Impact	A070102	High Capture Threshold	08/08/2025	08/08/2025
									A1502	Positioning Problem	08/08/2025	08/08/2025
									A23	Use of Device Problem	08/08/2025	08/08/2025
17470	2025	All other reportable incidents	47270	Cardiac resynchronization therapy implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A072201	High Impedance		
							F26	No Health Consequences or Impact	A072201	High Impedance		
17471	2025	All other reportable incidents	47263	Cardiac resynchronization therapy implantable pacemaker	CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	29/08/2025	29/08/2025
									A0908	Incorrect, Inadequate or Imprecise Result or Readings	29/08/2025	29/08/2025
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	29/08/2025	29/08/2025
									A0908	Incorrect, Inadequate or Imprecise Result or Readings	29/08/2025	29/08/2025
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	29/08/2025	29/08/2025
									A0908	Incorrect, Inadequate or Imprecise Result or Readings	29/08/2025	29/08/2025
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	29/08/2025	29/08/2025
									A0908	Incorrect, Inadequate or Imprecise Result or Readings	29/08/2025	29/08/2025
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	29/08/2025	29/08/2025
									A0908	Incorrect, Inadequate or Imprecise Result or Readings	29/08/2025	29/08/2025
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	29/08/2025	29/08/2025
									A0908	Incorrect, Inadequate or Imprecise Result or Readings	29/08/2025	29/08/2025
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	29/08/2025	29/08/2025
									A0908	Incorrect, Inadequate or Imprecise Result or Readings	29/08/2025	29/08/2025

							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	29/08/2025	29/08/2025
									A0908	Incorrect, Inadequate or Imprecise Result or Readings	29/08/2025	29/08/2025
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	29/08/2025	29/08/2025
									A0908	Incorrect, Inadequate or Imprecise Result or Readings	29/08/2025	29/08/2025
17472	2025	All other reportable incidents	47265	Dual-chamber implantable pacemaker, rate-responsive			E2403	No Clinical Signs, Symptoms or Conditions	A0908	Incorrect, Inadequate or Imprecise Result or Readings		
							F08	Hospitalization or Prolonged Hospitalization	A0908	Incorrect, Inadequate or Imprecise Result or Readings		
17473	2025	All other reportable incidents	63188	Left-ventricular wireless pacing system sensor/transmitter	CT145	Cardiovascular devices	E0623	Unspecified Heart Problem	A25	No Apparent Adverse Event		
							F26	No Health Consequences or Impact	A25	No Apparent Adverse Event		
					CT1577	Wireless monitoring system transmitters	E0623	Unspecified Heart Problem	A25	No Apparent Adverse Event		
							F26	No Health Consequences or Impact	A25	No Apparent Adverse Event		
					CT1684	Cardiac pacemakers and associated devices	E0623	Unspecified Heart Problem	A25	No Apparent Adverse Event		
							F26	No Health Consequences or Impact	A25	No Apparent Adverse Event		
					CT2727	Active implant devices/ancillaries	E0623	Unspecified Heart Problem	A25	No Apparent Adverse Event		
							F26	No Health Consequences or Impact	A25	No Apparent Adverse Event		
					CT2750	Implantable cardiac electrical stimulators and associated devices	E0623	Unspecified Heart Problem	A25	No Apparent Adverse Event		
							F26	No Health Consequences or Impact	A25	No Apparent Adverse Event		
					CT2885	Physiological monitoring/measuring system transmitters/receivers/transceivers	E0623	Unspecified Heart Problem	A25	No Apparent Adverse Event		
							F26	No Health Consequences or Impact	A25	No Apparent Adverse Event		
17474	2025	All other reportable incidents	47265	Dual-chamber implantable pacemaker, rate-responsive	CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
17475	2025	All other reportable incidents	35852	Single-chamber implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E060104	Bradycardia	A070103	Intermittent Capture		
							F26	No Health Consequences or Impact	A070103	Intermittent Capture		
17476	2025	All other reportable incidents	47205	Cardiac pulse generator programmer	CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A09	Output Problem		
									A13	Communication or Transmission Problem		
							F26	No Health Consequences or Impact	A09	Output Problem		
									A13	Communication or Transmission Problem		
17477	2025	All other reportable incidents	47205	Cardiac pulse generator programmer	CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A09	Output Problem		
									A13	Communication or Transmission Problem		
							F26	No Health Consequences or Impact	A09	Output Problem		
									A13	Communication or Transmission Problem		
17478	2025	All other reportable incidents	47263	Cardiac resynchronization therapy implantable pacemaker	CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A0908	Incorrect, Inadequate or Imprecise Result or Readings		
							F26	No Health Consequences or Impact	A0908	Incorrect, Inadequate or Imprecise Result or Readings		
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A0908	Incorrect, Inadequate or Imprecise Result or Readings		
							F26	No Health Consequences or Impact	A0908	Incorrect, Inadequate or Imprecise Result or Readings		
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A0908	Incorrect, Inadequate or Imprecise Result or Readings		
							F26	No Health Consequences or Impact	A0908	Incorrect, Inadequate or Imprecise Result or Readings		
17479	2025	All other reportable incidents	47265	Dual-chamber implantable pacemaker, rate-responsive	CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A071102	Failure to Interrogate		
							F26	No Health Consequences or Impact	A071102	Failure to Interrogate		
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A071102	Failure to Interrogate		
							F26	No Health Consequences or Impact	A071102	Failure to Interrogate		
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A071102	Failure to Interrogate		
							F26	No Health Consequences or Impact	A071102	Failure to Interrogate		
17480	2025	All other reportable incidents	47263	Cardiac resynchronization therapy implantable pacemaker	CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A070909	Oversensing		
									A0712	Pacing Problem		
									A090801	Signal Artifact/Noise		
							F1001	Absence of Treatment	A070909	Oversensing		
									A0712	Pacing Problem		
									A090801	Signal Artifact/Noise		
							F1908	Prolonged Surgery	A070909	Oversensing		
									A0712	Pacing Problem		
									A090801	Signal Artifact/Noise		
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A070909	Oversensing		
									A0712	Pacing Problem		
									A090801	Signal Artifact/Noise		
							F1001	Absence of Treatment	A070909	Oversensing		
									A0712	Pacing Problem		
									A090801	Signal Artifact/Noise		
							F1908	Prolonged Surgery	A070909	Oversensing		
									A0712	Pacing Problem		
									A090801	Signal Artifact/Noise		
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A070909	Oversensing		
									A0712	Pacing Problem		
									A090801	Signal Artifact/Noise		
							F1001	Absence of Treatment	A070909	Oversensing		
									A0712	Pacing Problem		
									A090801	Signal Artifact/Noise		
							F1908	Prolonged Surgery	A070909	Oversensing		
									A0712	Pacing Problem		
									A090801	Signal Artifact/Noise		
17481	2025	All other reportable incidents	47267	Single-chamber implantable pacemaker, rate-responsive			E0112	Dizziness	A0908	Incorrect, Inadequate or Imprecise Result or Readings	28/08/2025	28/08/2025
							F08	Hospitalization or Prolonged Hospitalization	A0908	Incorrect, Inadequate or Imprecise Result or Readings	28/08/2025	28/08/2025
							F1905	Device Revision or Replacement	A0908	Incorrect, Inadequate or Imprecise Result or Readings	28/08/2025	28/08/2025
17482	2025	All other reportable incidents	47270	Cardiac resynchronization therapy implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		

17483	2025	All other reportable incidents	35852	Single-chamber implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
									A0908	Incorrect, Inadequate or Imprecise Result or Readings		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
									A0908	Incorrect, Inadequate or Imprecise Result or Readings		
17484	2025	All other reportable incidents	35852	Single-chamber implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
17485	2025	All other reportable incidents	47205	Cardiac pulse generator programmer	CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A0708	Power Problem		
							F27	Problem Identified During Non-Clinical Procedure	A0708	Power Problem		
17486	2025	All other reportable incidents	47270	Cardiac resynchronization therapy implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E210401	Shock from Patient Lead(s)	A070910	Undersensing		
									A071302	Inappropriate/Inadequate Shock/Stimulation		
									A1301	Failure to Read Input Signal		
							F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A070910	Undersensing		
									A071302	Inappropriate/Inadequate Shock/Stimulation		
									A1301	Failure to Read Input Signal		
17487	2025	All other reportable incidents	37265	Dual-chamber implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E210401	Shock from Patient Lead(s)	A0712	Pacing Problem		
									A071302	Inappropriate/Inadequate Shock/Stimulation		
							E2311	Discomfort	A0712	Pacing Problem		
									A071302	Inappropriate/Inadequate Shock/Stimulation		
							E2330	Pain	A0712	Pacing Problem		
									A071302	Inappropriate/Inadequate Shock/Stimulation		
							F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A0712	Pacing Problem		
									A071302	Inappropriate/Inadequate Shock/Stimulation		
17488	2025	All other reportable incidents	47263	Cardiac resynchronization therapy implantable pacemaker	CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A070909	Oversensing		
									A072201	High Impedance		
									A090801	Signal Artifact/Noise		
							F26	No Health Consequences or Impact	A070909	Oversensing		
									A072201	High Impedance		
									A090801	Signal Artifact/Noise		
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A070909	Oversensing		
									A072201	High Impedance		
									A090801	Signal Artifact/Noise		
							F26	No Health Consequences or Impact	A070909	Oversensing		
									A072201	High Impedance		
									A090801	Signal Artifact/Noise		
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A070909	Oversensing		
									A072201	High Impedance		
									A090801	Signal Artifact/Noise		
							F26	No Health Consequences or Impact	A070909	Oversensing		
									A072201	High Impedance		
									A090801	Signal Artifact/Noise		
17489	2025	All other reportable incidents	63188	Left-ventricular wireless pacing system sensor/transmitter	CT145	Cardiovascular devices	E0623	Unspecified Heart Problem	A25	No Apparent Adverse Event		
							F26	No Health Consequences or Impact	A25	No Apparent Adverse Event		
					CT1577	Wireless monitoring system transmitters	E0623	Unspecified Heart Problem	A25	No Apparent Adverse Event		
							F26	No Health Consequences or Impact	A25	No Apparent Adverse Event		
					CT1684	Cardiac pacemakers and associated devices	E0623	Unspecified Heart Problem	A25	No Apparent Adverse Event		
							F26	No Health Consequences or Impact	A25	No Apparent Adverse Event		
					CT2727	Active implant devices/ancillaries	E0623	Unspecified Heart Problem	A25	No Apparent Adverse Event		
							F26	No Health Consequences or Impact	A25	No Apparent Adverse Event		
					CT2750	Implantable cardiac electrical stimulators and associated devices	E0623	Unspecified Heart Problem	A25	No Apparent Adverse Event		
							F26	No Health Consequences or Impact	A25	No Apparent Adverse Event		
					CT2885	Physiological monitoring/measuring system transmitters/receivers/transceivers	E0623	Unspecified Heart Problem	A25	No Apparent Adverse Event		
							F26	No Health Consequences or Impact	A25	No Apparent Adverse Event		
17490	2025	All other reportable incidents	36106	Implantable pacemaker analyser, intraoperative	CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A0902	Display or Visual Feedback Problem		
									A11	Computer Software Problem		
							F26	No Health Consequences or Impact	A0902	Display or Visual Feedback Problem		
									A11	Computer Software Problem		
17491	2025	All other reportable incidents	47270	Cardiac resynchronization therapy implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A05	Mechanical Problem	03/09/2025	03/09/2025
							F26	No Health Consequences or Impact	A05	Mechanical Problem	03/09/2025	03/09/2025
17492	2025	All other reportable incidents	47265	Dual-chamber implantable pacemaker, rate-responsive	CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A1107	Data Problem		
									A13	Communication or Transmission Problem		
							F26	No Health Consequences or Impact	A1107	Data Problem		
									A13	Communication or Transmission Problem		
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A1107	Data Problem		
									A13	Communication or Transmission Problem		
							F26	No Health Consequences or Impact	A1107	Data Problem		
									A13	Communication or Transmission Problem		
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A1107	Data Problem		
									A13	Communication or Transmission Problem		
							F26	No Health Consequences or Impact	A1107	Data Problem		
									A13	Communication or Transmission Problem		
17493	2025	All other reportable incidents	47267	Single-chamber implantable pacemaker, rate-responsive	CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A090805	Incorrect Measurement		
							F26	No Health Consequences or Impact	A090805	Incorrect Measurement		
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A090805	Incorrect Measurement		
							F26	No Health Consequences or Impact	A090805	Incorrect Measurement		
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A090805	Incorrect Measurement		
							F26	No Health Consequences or Impact	A090805	Incorrect Measurement		
17494	2025	All other reportable incidents	36106	Implantable pacemaker analyser, intraoperative	CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A13	Communication or Transmission Problem		
							F26	No Health Consequences or Impact	A13	Communication or Transmission Problem		
17495	2025	Unanticipated serious deterioration in state of health	60789	Intracardiac pacemaker	CT1685	Cardiac pacemakers	E0604	Cardiac Perforation	A24	No Apparent Device Problem		
							E0619	Pericardial Effusion	A24	No Apparent Device Problem		
							F08	Hospitalization or Prolonged Hospitalization	A24	No Apparent Device Problem		
							F12	Serious Injury/ Illness/ Impairment	A24	No Apparent Device Problem		
							F1203	Life Threatening Illness or Injury	A24	No Apparent Device Problem		
					CT257	Implantable cardiac pacemakers	E0604	Cardiac Perforation	A24	No Apparent Device Problem		
							E0619	Pericardial Effusion	A24	No Apparent Device Problem		
							F08	Hospitalization or Prolonged Hospitalization	A24	No Apparent Device Problem		

							F12	Serious Injury/ Illness/ Impairment	A24	No Apparent Device Problem		
							F1203	Life Threatening Illness or Injury	A24	No Apparent Device Problem		
					CT2818	Catheterization/Cannulation/Fluid path devices	E0604	Cardiac Perforation	A24	No Apparent Device Problem		
							E0619	Pericardial Effusion	A24	No Apparent Device Problem		
							F08	Hospitalization or Prolonged Hospitalization	A24	No Apparent Device Problem		
							F12	Serious Injury/ Illness/ Impairment	A24	No Apparent Device Problem		
							F1203	Life Threatening Illness or Injury	A24	No Apparent Device Problem		
					CT2866	Implantable cardiac electrical stimulators	E0604	Cardiac Perforation	A24	No Apparent Device Problem		
							E0619	Pericardial Effusion	A24	No Apparent Device Problem		
							F08	Hospitalization or Prolonged Hospitalization	A24	No Apparent Device Problem		
							F12	Serious Injury/ Illness/ Impairment	A24	No Apparent Device Problem		
							F1203	Life Threatening Illness or Injury	A24	No Apparent Device Problem		
17496	2025	All other reportable incidents	60789	Intracardiac pacemaker	CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A0701	Capturing Problem	09/09/2025	09/09/2025
							F26	No Health Consequences or Impact	A0701	Capturing Problem	09/09/2025	09/09/2025
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A0701	Capturing Problem	09/09/2025	09/09/2025
							F26	No Health Consequences or Impact	A0701	Capturing Problem	09/09/2025	09/09/2025
					CT2818	Catheterization/Cannulation/Fluid path devices	E2403	No Clinical Signs, Symptoms or Conditions	A0701	Capturing Problem	09/09/2025	09/09/2025
							F26	No Health Consequences or Impact	A0701	Capturing Problem	09/09/2025	09/09/2025
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A0701	Capturing Problem	09/09/2025	09/09/2025
							F26	No Health Consequences or Impact	A0701	Capturing Problem	09/09/2025	09/09/2025
17497	2025	All other reportable incidents	36106	Implantable pacemaker analyser, intraoperative	CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A11	Computer Software Problem		
									A13	Communication or Transmission Problem		
							F26	No Health Consequences or Impact	A11	Computer Software Problem		
									A13	Communication or Transmission Problem		
17498	2025	All other reportable incidents	47265	Dual-chamber implantable pacemaker, rate-responsive			E2403	No Clinical Signs, Symptoms or Conditions	A0712	Pacing Problem	08/08/2025	08/08/2025
									A12	Connection Problem	08/08/2025	08/08/2025
							F1908	Prolonged Surgery	A0712	Pacing Problem	08/08/2025	08/08/2025
									A12	Connection Problem	08/08/2025	08/08/2025
17499	2025	All other reportable incidents	47267	Single-chamber implantable pacemaker, rate-responsive			E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem	03/12/2025	03/12/2025
									A070504	Premature Discharge of Battery	03/12/2025	03/12/2025
									A0711	Interrogation Problem	03/12/2025	03/12/2025
							F08	Hospitalization or Prolonged Hospitalization	A0705	Battery Problem	03/12/2025	03/12/2025
									A070504	Premature Discharge of Battery	03/12/2025	03/12/2025
									A0711	Interrogation Problem	03/12/2025	03/12/2025
							F1905	Device Revision or Replacement	A0705	Battery Problem	03/12/2025	03/12/2025
									A070504	Premature Discharge of Battery	03/12/2025	03/12/2025
									A0711	Interrogation Problem	03/12/2025	03/12/2025
17500	2025	All other reportable incidents	17882	Physiologic-monitoring defibrillation system	CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A1603	Failure of Device to Self-Test		
							F27	Problem Identified During Non-Clinical Procedure	A1603	Failure of Device to Self-Test		
17501	2025	All other reportable incidents	47265	Dual-chamber implantable pacemaker, rate-responsive			E2403	No Clinical Signs, Symptoms or Conditions	A070909	Oversensing		
									A072201	High Impedance		
									A090801	Signal Artifact/Noise		
							F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A070909	Oversensing		
									A072201	High Impedance		
									A090801	Signal Artifact/Noise		
17502	2025	All other reportable incidents	17882	Physiologic-monitoring defibrillation system	CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A110401	Operating System Becomes Nonfunctional		
							F27	Problem Identified During Non-Clinical Procedure	A110401	Operating System Becomes Nonfunctional		
17503	2025	All other reportable incidents	37265	Dual-chamber implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A070909	Oversensing		
									A071302	Inappropriate/Inadequate Shock/Stimulation		
							F12	Serious Injury/ Illness/ Impairment	A070909	Oversensing		
									A071302	Inappropriate/Inadequate Shock/Stimulation		
17504	2025	Unanticipated serious deterioration in state of health	47265	Dual-chamber implantable pacemaker, rate-responsive			E0112	Dizziness	A0712	Pacing Problem		
							E060106	Heart Block	A0712	Pacing Problem		
							F08	Hospitalization or Prolonged Hospitalization	A0712	Pacing Problem		
							F12	Serious Injury/ Illness/ Impairment	A0712	Pacing Problem		
							F1905	Device Revision or Replacement	A0712	Pacing Problem		
17505	2025	All other reportable incidents	47263	Cardiac resynchronization therapy implantable pacemaker	CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
									A071102	Failure to Interrogate		
									A0712	Pacing Problem		
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery		
									A071102	Failure to Interrogate		
									A0712	Pacing Problem		
							F12	Serious Injury/ Illness/ Impairment	A070504	Premature Discharge of Battery		
									A071102	Failure to Interrogate		
									A0712	Pacing Problem		
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery		
									A071102	Failure to Interrogate		
									A0712	Pacing Problem		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
									A071102	Failure to Interrogate		
									A0712	Pacing Problem		
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
									A071102	Failure to Interrogate		
									A0712	Pacing Problem		
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery		
									A071102	Failure to Interrogate		
									A0712	Pacing Problem		
							F12	Serious Injury/ Illness/ Impairment	A070504	Premature Discharge of Battery		
									A071102	Failure to Interrogate		
									A0712	Pacing Problem		
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery		
									A071102	Failure to Interrogate		
									A0712	Pacing Problem		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
									A071102	Failure to Interrogate		
									A0712	Pacing Problem		
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
									A071102	Failure to Interrogate		

									A0712	Pacing Problem		
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery		
									A071102	Failure to Interrogate		
									A0712	Pacing Problem		
							F12	Serious Injury/ Illness/ Impairment	A070504	Premature Discharge of Battery		
									A071102	Failure to Interrogate		
									A0712	Pacing Problem		
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery		
									A071102	Failure to Interrogate		
									A0712	Pacing Problem		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
									A071102	Failure to Interrogate		
									A0712	Pacing Problem		
17506	2025	All other reportable incidents	36106	Implantable pacemaker analyser, intraoperative	CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A0709	Device Sensing Problem		
									A0902	Display or Visual Feedback Problem		
									A13	Communication or Transmission Problem		
							F23	Unexpected Medical Intervention	A0709	Device Sensing Problem		
									A0902	Display or Visual Feedback Problem		
									A13	Communication or Transmission Problem		
17507	2025	All other reportable incidents	36106	Implantable pacemaker analyser, intraoperative	CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A13	Communication or Transmission Problem		
							F26	No Health Consequences or Impact	A13	Communication or Transmission Problem		
17508	2025	All other reportable incidents	17882	Physiologic-monitoring defibrillation system	CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A0401	Break		
							F27	Problem Identified During Non-Clinical Procedure	A0401	Break		
17509	2025	All other reportable incidents	35852	Single-chamber implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E210401	Shock from Patient Lead(s)	A070909	Oversensing		
									A071302	Inappropriate/Inadequate Shock/Stimulation		
									A090801	Signal Artifact/Noise		
							F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A070909	Oversensing		
									A071302	Inappropriate/Inadequate Shock/Stimulation		
									A090801	Signal Artifact/Noise		
17510	2025	All other reportable incidents	35852	Single-chamber implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E210401	Shock from Patient Lead(s)	A070909	Oversensing		
									A071302	Inappropriate/Inadequate Shock/Stimulation		
									A1301	Failure to Read Input Signal		
							F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A070909	Oversensing		
									A071302	Inappropriate/Inadequate Shock/Stimulation		
									A1301	Failure to Read Input Signal		
17511	2025	All other reportable incidents	36106	Implantable pacemaker analyser, intraoperative	CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A1104	Computer Operating System Problem		
									A13	Communication or Transmission Problem		
							F26	No Health Consequences or Impact	A1104	Computer Operating System Problem		
									A13	Communication or Transmission Problem		
17512	2025	All other reportable incidents	35852	Single-chamber implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E210401	Shock from Patient Lead(s)	A0709	Device Sensing Problem		
									A070909	Oversensing		
									A071302	Inappropriate/Inadequate Shock/Stimulation		
									A1301	Failure to Read Input Signal		
									A23	Use of Device Problem		
							F08	Hospitalization or Prolonged Hospitalization	A0709	Device Sensing Problem		
									A070909	Oversensing		
									A071302	Inappropriate/Inadequate Shock/Stimulation		
									A1301	Failure to Read Input Signal		
									A23	Use of Device Problem		
							F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A0709	Device Sensing Problem		
									A070909	Oversensing		
									A071302	Inappropriate/Inadequate Shock/Stimulation		
									A1301	Failure to Read Input Signal		
									A23	Use of Device Problem		
17513	2025	All other reportable incidents	35852	Single-chamber implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A072201	High Impedance		
									A1301	Failure to Read Input Signal		
							F2203	Imaging Required	A072201	High Impedance		
									A1301	Failure to Read Input Signal		
							F26	No Health Consequences or Impact	A072201	High Impedance		
									A1301	Failure to Read Input Signal		
17514	2025	All other reportable incidents	47270	Cardiac resynchronization therapy implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A070909	Oversensing		
									A071204	No Pacing		
							F23	Unexpected Medical Intervention	A070909	Oversensing		
									A071204	No Pacing		
17515	2025	All other reportable incidents	35852	Single-chamber implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E210401	Shock from Patient Lead(s)	A0712	Pacing Problem		
									A071302	Inappropriate/Inadequate Shock/Stimulation		
							F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A0712	Pacing Problem		
									A071302	Inappropriate/Inadequate Shock/Stimulation		
17516	2025	All other reportable incidents	60789	Intracardiac pacemaker	CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A05	Mechanical Problem	12/09/2025	12/09/2025
									A1502	Positioning Problem	12/09/2025	12/09/2025
							F26	No Health Consequences or Impact	A05	Mechanical Problem	12/09/2025	12/09/2025
									A1502	Positioning Problem	12/09/2025	12/09/2025
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A05	Mechanical Problem	12/09/2025	12/09/2025
									A1502	Positioning Problem	12/09/2025	12/09/2025
							F26	No Health Consequences or Impact	A05	Mechanical Problem	12/09/2025	12/09/2025
									A1502	Positioning Problem	12/09/2025	12/09/2025
					CT2818	Catheterization/Cannulation/Fluid path devices	E2403	No Clinical Signs, Symptoms or Conditions	A05	Mechanical Problem	12/09/2025	12/09/2025
									A1502	Positioning Problem	12/09/2025	12/09/2025
							F26	No Health Consequences or Impact	A05	Mechanical Problem	12/09/2025	12/09/2025
									A1502	Positioning Problem	12/09/2025	12/09/2025
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A05	Mechanical Problem	12/09/2025	12/09/2025
									A1502	Positioning Problem	12/09/2025	12/09/2025
							F26	No Health Consequences or Impact	A05	Mechanical Problem	12/09/2025	12/09/2025
									A1502	Positioning Problem	12/09/2025	12/09/2025
17517	2025	All other reportable incidents	47265	Dual-chamber implantable pacemaker, rate-responsive	CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A071102	Failure to Interrogate		
							F26	No Health Consequences or Impact	A071102	Failure to Interrogate		
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A071102	Failure to Interrogate		
							F26	No Health Consequences or Impact	A071102	Failure to Interrogate		

					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A071102	Failure to Interrogate		
							F26	No Health Consequences or Impact	A071102	Failure to Interrogate		
17518	2025	All other reportable incidents	47265	Dual-chamber implantable pacemaker, rate-responsive			E2403	No Clinical Signs, Symptoms or Conditions	A071102	Failure to Interrogate	19/09/2025	19/09/2025
									A071204	No Pacing	19/09/2025	19/09/2025
							F1901	Additional Surgery	A071102	Failure to Interrogate	19/09/2025	19/09/2025
									A071204	No Pacing	19/09/2025	19/09/2025
17519	2025	All other reportable incidents	60789	Intracardiac pacemaker			E0503	Embolism/Embolus	A051201	Device Dislodged or Dislocated		
									A070101	Failure to Capture		
							F1901	Additional Surgery	A051201	Device Dislodged or Dislocated		
									A070101	Failure to Capture		
17520	2025	All other reportable incidents	47265	Dual-chamber implantable pacemaker, rate-responsive			E2403	No Clinical Signs, Symptoms or Conditions	A071102	Failure to Interrogate	18/09/2025	18/09/2025
							F1901	Additional Surgery	A071102	Failure to Interrogate	18/09/2025	18/09/2025
17521	2025	All other reportable incidents	47265	Dual-chamber implantable pacemaker, rate-responsive	CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A070909	Oversensing		
									A070910	Undersensing		
							F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A070909	Oversensing		
									A070910	Undersensing		
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A070909	Oversensing		
									A070910	Undersensing		
							F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A070909	Oversensing		
									A070910	Undersensing		
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A070909	Oversensing		
									A070910	Undersensing		
							F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A070909	Oversensing		
									A070910	Undersensing		
17522	2025	All other reportable incidents	47265	Dual-chamber implantable pacemaker, rate-responsive			E2403	No Clinical Signs, Symptoms or Conditions	A0908	Incorrect, Inadequate or Imprecise Result or Readings	13/09/2025	13/09/2025
							F08	Hospitalization or Prolonged Hospitalization	A0908	Incorrect, Inadequate or Imprecise Result or Readings	13/09/2025	13/09/2025
							F1905	Device Revision or Replacement	A0908	Incorrect, Inadequate or Imprecise Result or Readings	13/09/2025	13/09/2025
17523	2025		63188	Left-ventricular wireless pacing system sensor/transmitter	CT145	Cardiovascular devices	E0102	Brain Injury	A0705	Battery Problem		
									A070603	Failure to Charge		
							E0119	Loss of Consciousness	A0705	Battery Problem		
									A070603	Failure to Charge		
							E0726	Hypoxia	A0705	Battery Problem		
									A070603	Failure to Charge		
							F02	Death	A0705	Battery Problem		
									A070603	Failure to Charge		
					CT1577	Wireless monitoring system transmitters	E0102	Brain Injury	A0705	Battery Problem		
									A070603	Failure to Charge		
							E0119	Loss of Consciousness	A0705	Battery Problem		
									A070603	Failure to Charge		
							E0726	Hypoxia	A0705	Battery Problem		
									A070603	Failure to Charge		
							F02	Death	A0705	Battery Problem		
									A070603	Failure to Charge		
					CT1684	Cardiac pacemakers and associated devices	E0102	Brain Injury	A0705	Battery Problem		
									A070603	Failure to Charge		
							E0119	Loss of Consciousness	A0705	Battery Problem		
									A070603	Failure to Charge		
							E0726	Hypoxia	A0705	Battery Problem		
									A070603	Failure to Charge		
							F02	Death	A0705	Battery Problem		
									A070603	Failure to Charge		
					CT2727	Active implant devices/ancillaries	E0102	Brain Injury	A0705	Battery Problem		
									A070603	Failure to Charge		
							E0119	Loss of Consciousness	A0705	Battery Problem		
									A070603	Failure to Charge		
							E0726	Hypoxia	A0705	Battery Problem		
									A070603	Failure to Charge		
							F02	Death	A0705	Battery Problem		
									A070603	Failure to Charge		
					CT2750	Implantable cardiac electrical stimulators and associated devices	E0102	Brain Injury	A0705	Battery Problem		
									A070603	Failure to Charge		
							E0119	Loss of Consciousness	A0705	Battery Problem		
									A070603	Failure to Charge		
							E0726	Hypoxia	A0705	Battery Problem		
									A070603	Failure to Charge		
							F02	Death	A0705	Battery Problem		
									A070603	Failure to Charge		
					CT2885	Physiological monitoring/measuring system transmitters/receivers/transceivers	E0102	Brain Injury	A0705	Battery Problem		
									A070603	Failure to Charge		
							E0119	Loss of Consciousness	A0705	Battery Problem		
									A070603	Failure to Charge		
							E0726	Hypoxia	A0705	Battery Problem		
									A070603	Failure to Charge		
							F02	Death	A0705	Battery Problem		
									A070603	Failure to Charge		
17524	2025	All other reportable incidents	37265	Dual-chamber implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A072201	High Impedance		
							F26	No Health Consequences or Impact	A072201	High Impedance		
17525	2025	All other reportable incidents	47270	Cardiac resynchronization therapy implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A070909	Oversensing		
									A1107	Data Problem		
									A23	Use of Device Problem		
							F08	Hospitalization or Prolonged Hospitalization	A070909	Oversensing		
									A1107	Data Problem		
									A23	Use of Device Problem		
							F2203	Imaging Required	A070909	Oversensing		
									A1107	Data Problem		

									A23	Use of Device Problem		
							F23	Unexpected Medical Intervention	A070909	Oversensing		
									A1107	Data Problem		
									A23	Use of Device Problem		
17526	2025	All other reportable incidents	47265	Dual-chamber implantable pacemaker, rate-responsive	CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	26/09/2025	26/09/2025
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	26/09/2025	26/09/2025
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	26/09/2025	26/09/2025
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery	26/09/2025	26/09/2025
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	26/09/2025	26/09/2025
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	26/09/2025	26/09/2025
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	26/09/2025	26/09/2025
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery	26/09/2025	26/09/2025
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	26/09/2025	26/09/2025
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	26/09/2025	26/09/2025
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	26/09/2025	26/09/2025
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery	26/09/2025	26/09/2025
17527	2025	All other reportable incidents	17882	Physiologic-monitoring defibrillation system	CT1684	Cardiac pacemakers and associated devices	E0602	Cardiac Arrest	A07	Electrical /Electronic Property Problem		
							F05	Delay to Treatment/ Therapy	A07	Electrical /Electronic Property Problem		
17528	2025	All other reportable incidents	47267	Single-chamber implantable pacemaker, rate-responsive			E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	31/10/2025	31/10/2025
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	31/10/2025	31/10/2025
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	31/10/2025	31/10/2025
17529	2025	All other reportable incidents	47270	Cardiac resynchronization therapy implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E210401	Shock from Patient Lead(s)	A070909	Oversensing		
									A071302	Inappropriate/Inadequate Shock/Stimulation		
									A0908	Incorrect, Inadequate or Imprecise Result or Readings		
							F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A070909	Oversensing		
									A071302	Inappropriate/Inadequate Shock/Stimulation		
									A0908	Incorrect, Inadequate or Imprecise Result or Readings		
17530	2025	All other reportable incidents	47265	Dual-chamber implantable pacemaker, rate-responsive	CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A160503	Inappropriate or Unexpected Reset		
							F26	No Health Consequences or Impact	A160503	Inappropriate or Unexpected Reset		
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A160503	Inappropriate or Unexpected Reset		
							F26	No Health Consequences or Impact	A160503	Inappropriate or Unexpected Reset		
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A160503	Inappropriate or Unexpected Reset		
							F26	No Health Consequences or Impact	A160503	Inappropriate or Unexpected Reset		
17531	2025	All other reportable incidents	47265	Dual-chamber implantable pacemaker, rate-responsive	CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A0908	Incorrect, Inadequate or Imprecise Result or Readings	23/09/2025	23/09/2025
							F08	Hospitalization or Prolonged Hospitalization	A0908	Incorrect, Inadequate or Imprecise Result or Readings	23/09/2025	23/09/2025
							F1905	Device Revision or Replacement	A0908	Incorrect, Inadequate or Imprecise Result or Readings	23/09/2025	23/09/2025
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A0908	Incorrect, Inadequate or Imprecise Result or Readings	23/09/2025	23/09/2025
							F08	Hospitalization or Prolonged Hospitalization	A0908	Incorrect, Inadequate or Imprecise Result or Readings	23/09/2025	23/09/2025
							F1905	Device Revision or Replacement	A0908	Incorrect, Inadequate or Imprecise Result or Readings	23/09/2025	23/09/2025
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A0908	Incorrect, Inadequate or Imprecise Result or Readings	23/09/2025	23/09/2025
							F08	Hospitalization or Prolonged Hospitalization	A0908	Incorrect, Inadequate or Imprecise Result or Readings	23/09/2025	23/09/2025
							F1905	Device Revision or Replacement	A0908	Incorrect, Inadequate or Imprecise Result or Readings	23/09/2025	23/09/2025
17532	2025	All other reportable incidents	47265	Dual-chamber implantable pacemaker, rate-responsive			E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem		
							F26	No Health Consequences or Impact	A0705	Battery Problem		
17533	2025	All other reportable incidents	35852	Single-chamber implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E210401	Shock from Patient Lead(s)	A070909	Oversensing		
									A071302	Inappropriate/Inadequate Shock/Stimulation		
							F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A070909	Oversensing		
									A071302	Inappropriate/Inadequate Shock/Stimulation		
17534	2025	All other reportable incidents	35852	Single-chamber implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E210401	Shock from Patient Lead(s)	A0709	Device Sensing Problem		
									A071302	Inappropriate/Inadequate Shock/Stimulation		
							F1203	Life Threatening Illness or Injury	A0709	Device Sensing Problem		
									A071302	Inappropriate/Inadequate Shock/Stimulation		
17535	2025	All other reportable incidents	47265	Dual-chamber implantable pacemaker, rate-responsive			E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
17536	2025	All other reportable incidents	36106	Implantable pacemaker analyser, intraoperative	CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A11	Computer Software Problem		
							F23	Unexpected Medical Intervention	A11	Computer Software Problem		
17537	2025	All other reportable incidents	47265	Dual-chamber implantable pacemaker, rate-responsive	CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A070102	High Capture Threshold	08/10/2025	08/10/2025
									A0705	Battery Problem	08/10/2025	08/10/2025
									A070504	Premature Discharge of Battery	08/10/2025	08/10/2025
									A072202	Low Impedance	08/10/2025	08/10/2025
							F08	Hospitalization or Prolonged Hospitalization	A070102	High Capture Threshold	08/10/2025	08/10/2025
									A0705	Battery Problem	08/10/2025	08/10/2025
									A070504	Premature Discharge of Battery	08/10/2025	08/10/2025
									A072202	Low Impedance	08/10/2025	08/10/2025
							F1905	Device Revision or Replacement	A070102	High Capture Threshold	08/10/2025	08/10/2025
									A0705	Battery Problem	08/10/2025	08/10/2025
									A070504	Premature Discharge of Battery	08/10/2025	08/10/2025
									A072202	Low Impedance	08/10/2025	08/10/2025
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A070102	High Capture Threshold	08/10/2025	08/10/2025
									A0705	Battery Problem	08/10/2025	08/10/2025
									A070504	Premature Discharge of Battery	08/10/2025	08/10/2025
									A072202	Low Impedance	08/10/2025	08/10/2025
							F08	Hospitalization or Prolonged Hospitalization	A070102	High Capture Threshold	08/10/2025	08/10/2025
									A0705	Battery Problem	08/10/2025	08/10/2025
									A070504	Premature Discharge of Battery	08/10/2025	08/10/2025
									A072202	Low Impedance	08/10/2025	08/10/2025
							F1905	Device Revision or Replacement	A070102	High Capture Threshold	08/10/2025	08/10/2025
									A0705	Battery Problem	08/10/2025	08/10/2025
									A070504	Premature Discharge of Battery	08/10/2025	08/10/2025

									A072202	Low Impedance	08/10/2025	08/10/2025
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A070102	High Capture Threshold	08/10/2025	08/10/2025
									A0705	Battery Problem	08/10/2025	08/10/2025
									A070504	Premature Discharge of Battery	08/10/2025	08/10/2025
									A072202	Low Impedance	08/10/2025	08/10/2025
							F08	Hospitalization or Prolonged Hospitalization	A070102	High Capture Threshold	08/10/2025	08/10/2025
									A0705	Battery Problem	08/10/2025	08/10/2025
									A070504	Premature Discharge of Battery	08/10/2025	08/10/2025
									A072202	Low Impedance	08/10/2025	08/10/2025
							F1905	Device Revision or Replacement	A070102	High Capture Threshold	08/10/2025	08/10/2025
									A0705	Battery Problem	08/10/2025	08/10/2025
									A070504	Premature Discharge of Battery	08/10/2025	08/10/2025
									A072202	Low Impedance	08/10/2025	08/10/2025
17538	2025	All other reportable incidents	35852	Single-chamber implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E210401	Shock from Patient Lead(s)	A070909	Oversensing		
									A071302	Inappropriate/Inadequate Shock/Stimulation		
									A090801	Signal Artifact/Noise		
									A1301	Failure to Read Input Signal		
							F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A070909	Oversensing		
									A071302	Inappropriate/Inadequate Shock/Stimulation		
									A090801	Signal Artifact/Noise		
									A1301	Failure to Read Input Signal		
							F2203	Imaging Required	A070909	Oversensing		
									A071302	Inappropriate/Inadequate Shock/Stimulation		
									A090801	Signal Artifact/Noise		
									A1301	Failure to Read Input Signal		
17539	2025	All other reportable incidents	47265	Dual-chamber implantable pacemaker, rate-responsive			E2403	No Clinical Signs, Symptoms or Conditions	A072201	High Impedance		
									A090801	Signal Artifact/Noise		
							F26	No Health Consequences or Impact	A072201	High Impedance		
									A090801	Signal Artifact/Noise		
17540	2025	Death	35852	Single-chamber implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem	18/08/2025	18/08/2025
									A0908	Incorrect, Inadequate or Imprecise Result or Readings	18/08/2025	18/08/2025
							F02	Death	A0705	Battery Problem	18/08/2025	18/08/2025
									A0908	Incorrect, Inadequate or Imprecise Result or Readings	18/08/2025	18/08/2025
17541	2025	All other reportable incidents	47267	Single-chamber implantable pacemaker, rate-responsive			E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
17542	2025		47270	Cardiac resynchronization therapy implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	18/09/2025	
							F1901	Additional Surgery	A070504	Premature Discharge of Battery	18/09/2025	
							F1903	Device Explantation	A070504	Premature Discharge of Battery	18/09/2025	
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	18/09/2025	
17543	2025	All other reportable incidents	17882	Physiologic-monitoring defibrillation system	CT1684	Cardiac pacemakers and associated devices	E060110	Ventricular Fibrillation	A071301	Failure to Deliver Shock/Stimulation		
							F02	Death	A071301	Failure to Deliver Shock/Stimulation		
17544	2025	All other reportable incidents	47205	Cardiac pulse generator programmer	CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A11	Computer Software Problem		
									A2202	Device Difficult to Program or Calibrate		
							F26	No Health Consequences or Impact	A11	Computer Software Problem		
									A2202	Device Difficult to Program or Calibrate		
17545	2025	All other reportable incidents	47205	Cardiac pulse generator programmer	CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A11	Computer Software Problem		
									A1104	Computer Operating System Problem		
									A1107	Data Problem		
							F26	No Health Consequences or Impact	A11	Computer Software Problem		
									A1104	Computer Operating System Problem		
									A1107	Data Problem		
17546	2025	All other reportable incidents	47265	Dual-chamber implantable pacemaker, rate-responsive	CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem		
							F26	No Health Consequences or Impact	A0705	Battery Problem		
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem		
							F26	No Health Consequences or Impact	A0705	Battery Problem		
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem		
							F26	No Health Consequences or Impact	A0705	Battery Problem		
17547	2025	All other reportable incidents	47265	Dual-chamber implantable pacemaker, rate-responsive	CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem		
							F26	No Health Consequences or Impact	A0705	Battery Problem		
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem		
							F26	No Health Consequences or Impact	A0705	Battery Problem		
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem		
							F26	No Health Consequences or Impact	A0705	Battery Problem		
17548	2025	All other reportable incidents	47265	Dual-chamber implantable pacemaker, rate-responsive	CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem		
							F26	No Health Consequences or Impact	A0705	Battery Problem		
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem		
							F26	No Health Consequences or Impact	A0705	Battery Problem		
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem		
							F26	No Health Consequences or Impact	A0705	Battery Problem		
17549	2025	All other reportable incidents	37265	Dual-chamber implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A070101	Failure to Capture		
									A070910	Undersensing		
									A0712	Pacing Problem		
							F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A070101	Failure to Capture		
									A070910	Undersensing		
									A0712	Pacing Problem		
							F15	Recognized Device or Procedural Complication	A070101	Failure to Capture		
									A070910	Undersensing		
									A0712	Pacing Problem		
17550	2025	All other reportable incidents	47270	Cardiac resynchronization therapy implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A1107	Data Problem		
							F26	No Health Consequences or Impact	A1107	Data Problem		
17551	2025		35852	Single-chamber implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
									A160602	Premature End-of-Life Indicator		
							F1905	Device Revision or Replacement	A0705	Battery Problem		

									A070504	Premature Discharge of Battery		
									A160602	Premature End-of-Life Indicator		
17552	2025	All other reportable incidents	60789	Intracardiac pacemaker			E2403	No Clinical Signs, Symptoms or Conditions	A051201	Device Dislodged or Dislocated		
									A070102	High Capture Threshold		
									A0722	Impedance Problem		
									A072201	High Impedance		
							F26	No Health Consequences or Impact	A051201	Device Dislodged or Dislocated		
									A070102	High Capture Threshold		
									A0722	Impedance Problem		
									A072201	High Impedance		
17553	2025	All other reportable incidents	36106	Implantable pacemaker analyser, intraoperative	CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A11	Computer Software Problem		
							F26	No Health Consequences or Impact	A11	Computer Software Problem		
17554	2025	All other reportable incidents	37265	Dual-chamber implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	09/09/2025	09/09/2025
							F1901	Additional Surgery	A070504	Premature Discharge of Battery	09/09/2025	09/09/2025
17555	2025	All other reportable incidents	47205	Cardiac pulse generator programmer	CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A09	Output Problem		
									A2202	Device Difficult to Program or Calibrate		
							F26	No Health Consequences or Impact	A09	Output Problem		
									A2202	Device Difficult to Program or Calibrate		
17556	2025		47265	Dual-chamber implantable pacemaker, rate-responsive			E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	11/07/2025	11/07/2025
							F1901	Additional Surgery	A24	No Apparent Device Problem	11/07/2025	11/07/2025
17557	2025	All other reportable incidents	47265	Dual-chamber implantable pacemaker, rate-responsive			E2403	No Clinical Signs, Symptoms or Conditions	A0908	Incorrect, Inadequate or Imprecise Result or Readings		
							F26	No Health Consequences or Impact	A0908	Incorrect, Inadequate or Imprecise Result or Readings		
17558	2025	All other reportable incidents	47265	Dual-chamber implantable pacemaker, rate-responsive			E2403	No Clinical Signs, Symptoms or Conditions	A0908	Incorrect, Inadequate or Imprecise Result or Readings	11/03/2025	11/03/2025
							F08	Hospitalization or Prolonged Hospitalization	A0908	Incorrect, Inadequate or Imprecise Result or Readings	11/03/2025	11/03/2025
							F1905	Device Revision or Replacement	A0908	Incorrect, Inadequate or Imprecise Result or Readings	11/03/2025	11/03/2025
17559	2025	All other reportable incidents	47270	Cardiac resynchronization therapy implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
17560	2025	All other reportable incidents	47267	Single-chamber implantable pacemaker, rate-responsive			E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem		
									A070909	Oversensing		
									A072201	High Impedance		
									A090801	Signal Artifact/Noise		
									A1301	Failure to Read Input Signal		
							F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A0705	Battery Problem		
									A070909	Oversensing		
									A072201	High Impedance		
									A090801	Signal Artifact/Noise		
									A1301	Failure to Read Input Signal		
17561	2025	All other reportable incidents	47265	Dual-chamber implantable pacemaker, rate-responsive			E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
17562	2025	All other reportable incidents	47270	Cardiac resynchronization therapy implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
									A0908	Incorrect, Inadequate or Imprecise Result or Readings		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
									A0908	Incorrect, Inadequate or Imprecise Result or Readings		
17563	2025	All other reportable incidents	47265	Dual-chamber implantable pacemaker, rate-responsive			E2103	Device Overstimulation of Tissue	A12	Connection Problem		
							F2203	Imaging Required	A12	Connection Problem		
17564	2025	All other reportable incidents	47270	Cardiac resynchronization therapy implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E210401	Shock from Patient Lead(s)	A070909	Oversensing		
									A071302	Inappropriate/Inadequate Shock/Stimulation		
							F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A070909	Oversensing		
									A071302	Inappropriate/Inadequate Shock/Stimulation		
17565	2025	All other reportable incidents	35852	Single-chamber implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
17566	2025	All other reportable incidents	47265	Dual-chamber implantable pacemaker, rate-responsive	CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
17567	2025	All other reportable incidents	47265	Dual-chamber implantable pacemaker, rate-responsive	CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery		
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery		
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery		
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery		
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery		
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery		
17568	2025	All other reportable incidents	63188	Left-ventricular wireless pacing system sensor/transmitter			E0623	Unspecified Heart Problem	A26	Insufficient Device Problem Information	03/10/2025	03/10/2025
							F15	Recognized Device or Procedural Complication	A26	Insufficient Device Problem Information	03/10/2025	03/10/2025
17569	2025	All other reportable incidents	47265	Dual-chamber implantable pacemaker, rate-responsive			E2403	No Clinical Signs, Symptoms or Conditions	A070101	Failure to Capture		
							F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A070101	Failure to Capture		
17570	2025	All other reportable incidents	47265	Dual-chamber implantable pacemaker, rate-responsive			E2403	No Clinical Signs, Symptoms or Conditions	A070909	Oversensing		
									A090801	Signal Artifact/Noise		
							F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A070909	Oversensing		
									A090801	Signal Artifact/Noise		
17571	2025	All other reportable incidents	36106	Implantable pacemaker analyser, intraoperative	CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A11	Computer Software Problem		
							F23	Unexpected Medical Intervention	A11	Computer Software Problem		
17572	2025	All other reportable incidents	36106	Implantable pacemaker analyser, intraoperative	CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A11	Computer Software Problem		
							F23	Unexpected Medical Intervention	A11	Computer Software Problem		
17573	2025	All other reportable incidents	47265	Dual-chamber implantable pacemaker, rate-responsive			E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem	29/10/2025	29/10/2025
									A070504	Premature Discharge of Battery	29/10/2025	29/10/2025
							F08	Hospitalization or Prolonged Hospitalization	A0705	Battery Problem	29/10/2025	29/10/2025
									A070504	Premature Discharge of Battery	29/10/2025	29/10/2025

							F1905	Device Revision or Replacement	A0705	Battery Problem	29/10/2025	29/10/2025
									A070504	Premature Discharge of Battery	29/10/2025	29/10/2025
17574	2025	All other reportable incidents	47265	Dual-chamber implantable pacemaker, rate-responsive	CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
17575	2025	All other reportable incidents	35224	External pacemaker, epicardial pacing	CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A08	Calibration Problem		
							F27	Problem Identified During Non-Clinical Procedure	A08	Calibration Problem		
17576	2025	Unanticipated serious deterioration in state of health	60789	Intracardiac pacemaker	CT1685	Cardiac pacemakers	E0506	Hemorrhage/Blood Loss/Bleeding	A24	No Apparent Device Problem		
							F08	Hospitalization or Prolonged Hospitalization	A24	No Apparent Device Problem		
							F12	Serious Injury/ Illness/ Impairment	A24	No Apparent Device Problem		
							F1901	Additional Surgery	A24	No Apparent Device Problem		
					CT257	Implantable cardiac pacemakers	E0506	Hemorrhage/Blood Loss/Bleeding	A24	No Apparent Device Problem		
							F08	Hospitalization or Prolonged Hospitalization	A24	No Apparent Device Problem		
							F12	Serious Injury/ Illness/ Impairment	A24	No Apparent Device Problem		
							F1901	Additional Surgery	A24	No Apparent Device Problem		
					CT2818	Catheterization/Cannulation/Fluid path devices	E0506	Hemorrhage/Blood Loss/Bleeding	A24	No Apparent Device Problem		
							F08	Hospitalization or Prolonged Hospitalization	A24	No Apparent Device Problem		
							F12	Serious Injury/ Illness/ Impairment	A24	No Apparent Device Problem		
							F1901	Additional Surgery	A24	No Apparent Device Problem		
					CT2866	Implantable cardiac electrical stimulators	E0506	Hemorrhage/Blood Loss/Bleeding	A24	No Apparent Device Problem		
							F08	Hospitalization or Prolonged Hospitalization	A24	No Apparent Device Problem		
							F12	Serious Injury/ Illness/ Impairment	A24	No Apparent Device Problem		
							F1901	Additional Surgery	A24	No Apparent Device Problem		
17577	2025	All other reportable incidents	36106	Implantable pacemaker analyser, intraoperative	CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A11	Computer Software Problem		
							F23	Unexpected Medical Intervention	A11	Computer Software Problem		
17578	2025	All other reportable incidents	47265	Dual-chamber implantable pacemaker, rate-responsive	CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A070909	Oversensing	02/10/2025	02/10/2025
									A0712	Pacing Problem	02/10/2025	02/10/2025
									A090801	Signal Artifact/Noise	02/10/2025	02/10/2025
							F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A070909	Oversensing	02/10/2025	02/10/2025
									A0712	Pacing Problem	02/10/2025	02/10/2025
									A090801	Signal Artifact/Noise	02/10/2025	02/10/2025
							F1001	Absence of Treatment	A070909	Oversensing	02/10/2025	02/10/2025
									A0712	Pacing Problem	02/10/2025	02/10/2025
									A090801	Signal Artifact/Noise	02/10/2025	02/10/2025
							F1908	Prolonged Surgery	A070909	Oversensing	02/10/2025	02/10/2025
									A0712	Pacing Problem	02/10/2025	02/10/2025
									A090801	Signal Artifact/Noise	02/10/2025	02/10/2025
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A070909	Oversensing	02/10/2025	02/10/2025
									A0712	Pacing Problem	02/10/2025	02/10/2025
									A090801	Signal Artifact/Noise	02/10/2025	02/10/2025
							F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A070909	Oversensing	02/10/2025	02/10/2025
									A0712	Pacing Problem	02/10/2025	02/10/2025
									A090801	Signal Artifact/Noise	02/10/2025	02/10/2025
							F1001	Absence of Treatment	A070909	Oversensing	02/10/2025	02/10/2025
									A0712	Pacing Problem	02/10/2025	02/10/2025
									A090801	Signal Artifact/Noise	02/10/2025	02/10/2025
							F1908	Prolonged Surgery	A070909	Oversensing	02/10/2025	02/10/2025
									A0712	Pacing Problem	02/10/2025	02/10/2025
									A090801	Signal Artifact/Noise	02/10/2025	02/10/2025
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A070909	Oversensing	02/10/2025	02/10/2025
									A0712	Pacing Problem	02/10/2025	02/10/2025
									A090801	Signal Artifact/Noise	02/10/2025	02/10/2025
							F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A070909	Oversensing	02/10/2025	02/10/2025
									A0712	Pacing Problem	02/10/2025	02/10/2025
									A090801	Signal Artifact/Noise	02/10/2025	02/10/2025
							F1001	Absence of Treatment	A070909	Oversensing	02/10/2025	02/10/2025
									A0712	Pacing Problem	02/10/2025	02/10/2025
									A090801	Signal Artifact/Noise	02/10/2025	02/10/2025
							F1908	Prolonged Surgery	A070909	Oversensing	02/10/2025	02/10/2025
									A0712	Pacing Problem	02/10/2025	02/10/2025
									A090801	Signal Artifact/Noise	02/10/2025	02/10/2025
17579	2025	All other reportable incidents	36106	Implantable pacemaker analyser, intraoperative	CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A13	Communication or Transmission Problem		
							F26	No Health Consequences or Impact	A13	Communication or Transmission Problem		
17580	2025	All other reportable incidents	47265	Dual-chamber implantable pacemaker, rate-responsive			E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
17581	2025	All other reportable incidents	35852	Single-chamber implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
17582	2025	All other reportable incidents	47265	Dual-chamber implantable pacemaker, rate-responsive			E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	07/11/2025	07/11/2025
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	07/11/2025	07/11/2025
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	07/11/2025	07/11/2025
17583	2025	All other reportable incidents	47265	Dual-chamber implantable pacemaker, rate-responsive			E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
17584	2025	All other reportable incidents	47265	Dual-chamber implantable pacemaker, rate-responsive	CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
17585	2025	All other reportable incidents	47265	Dual-chamber implantable pacemaker, rate-responsive			E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
17586	2025	All other reportable incidents	47267	Single-chamber implantable pacemaker, rate-responsive			E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
17587	2025	Death	17882	Physiologic-monitoring defibrillation system	CT1684	Cardiac pacemakers and associated devices	E0601	Arrhythmia	A07	Electrical /Electronic Property Problem		

							F02	Death	A07	Electrical /Electronic Property Problem		
17588	2025	Unanticipated serious deterioration in state of health	60789	Intracardiac pacemaker	CT1685	Cardiac pacemakers	E0602	Cardiac Arrest	A24	No Apparent Device Problem		
							E0611	Heart Failure/Congestive Heart Failure	A24	No Apparent Device Problem		
							E0717	Dyspnea	A24	No Apparent Device Problem		
							F02	Death	A24	No Apparent Device Problem		
					CT257	Implantable cardiac pacemakers	E0602	Cardiac Arrest	A24	No Apparent Device Problem		
							E0611	Heart Failure/Congestive Heart Failure	A24	No Apparent Device Problem		
							E0717	Dyspnea	A24	No Apparent Device Problem		
							F02	Death	A24	No Apparent Device Problem		
					CT2818	Catheterization/Cannulation/Fluid path devices	E0602	Cardiac Arrest	A24	No Apparent Device Problem		
							E0611	Heart Failure/Congestive Heart Failure	A24	No Apparent Device Problem		
							E0717	Dyspnea	A24	No Apparent Device Problem		
							F02	Death	A24	No Apparent Device Problem		
					CT2866	Implantable cardiac electrical stimulators	E0602	Cardiac Arrest	A24	No Apparent Device Problem		
							E0611	Heart Failure/Congestive Heart Failure	A24	No Apparent Device Problem		
							E0717	Dyspnea	A24	No Apparent Device Problem		
							F02	Death	A24	No Apparent Device Problem		
17589	2025	All other reportable incidents	35852	Single-chamber implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A0711	Interrogation Problem		
							F26	No Health Consequences or Impact	A0711	Interrogation Problem		
17590	2025	All other reportable incidents	47267	Single-chamber implantable pacemaker, rate-responsive			E2403	No Clinical Signs, Symptoms or Conditions	A0908	Incorrect, Inadequate or Imprecise Result or Readings	13/10/2025	13/10/2025
							F08	Hospitalization or Prolonged Hospitalization	A0908	Incorrect, Inadequate or Imprecise Result or Readings	13/10/2025	13/10/2025
							F1905	Device Revision or Replacement	A0908	Incorrect, Inadequate or Imprecise Result or Readings	13/10/2025	13/10/2025
17591	2025	All other reportable incidents	47265	Dual-chamber implantable pacemaker, rate-responsive			E0112	Dizziness	A070504	Premature Discharge of Battery		
									A071102	Failure to Interrogate		
									A071204	No Pacing		
							E0601	Arrhythmia	A070504	Premature Discharge of Battery		
									A071102	Failure to Interrogate		
									A071204	No Pacing		
							F1901	Additional Surgery	A070504	Premature Discharge of Battery		
									A071102	Failure to Interrogate		
									A071204	No Pacing		
17592	2025	All other reportable incidents	47265	Dual-chamber implantable pacemaker, rate-responsive			E2403	No Clinical Signs, Symptoms or Conditions	A0908	Incorrect, Inadequate or Imprecise Result or Readings	10/10/2025	10/10/2025
							F08	Hospitalization or Prolonged Hospitalization	A0908	Incorrect, Inadequate or Imprecise Result or Readings	10/10/2025	10/10/2025
							F1905	Device Revision or Replacement	A0908	Incorrect, Inadequate or Imprecise Result or Readings	10/10/2025	10/10/2025
17593	2025	All other reportable incidents	37265	Dual-chamber implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A160601	Premature Elective Replacement Indicator		
							F26	No Health Consequences or Impact	A160601	Premature Elective Replacement Indicator		
17594	2025		47265	Dual-chamber implantable pacemaker, rate-responsive	CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A02	Manufacturing, Packaging or Shipping Problem		
							F26	No Health Consequences or Impact	A02	Manufacturing, Packaging or Shipping Problem		
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A02	Manufacturing, Packaging or Shipping Problem		
							F26	No Health Consequences or Impact	A02	Manufacturing, Packaging or Shipping Problem		
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A02	Manufacturing, Packaging or Shipping Problem		
							F26	No Health Consequences or Impact	A02	Manufacturing, Packaging or Shipping Problem		
17595	2025	All other reportable incidents	47265	Dual-chamber implantable pacemaker, rate-responsive			E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	01/01/2025	01/01/2025
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	01/01/2025	01/01/2025
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	01/01/2025	01/01/2025
17596	2025	All other reportable incidents	47265	Dual-chamber implantable pacemaker, rate-responsive			E2403	No Clinical Signs, Symptoms or Conditions	A0908	Incorrect, Inadequate or Imprecise Result or Readings	06/10/2025	06/10/2025
							F08	Hospitalization or Prolonged Hospitalization	A0908	Incorrect, Inadequate or Imprecise Result or Readings	06/10/2025	06/10/2025
							F1905	Device Revision or Replacement	A0908	Incorrect, Inadequate or Imprecise Result or Readings	06/10/2025	06/10/2025
17597	2025	All other reportable incidents	47265	Dual-chamber implantable pacemaker, rate-responsive			E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	30/10/2025	30/10/2025
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	30/10/2025	30/10/2025
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	30/10/2025	30/10/2025
17598	2025	All other reportable incidents	47265	Dual-chamber implantable pacemaker, rate-responsive	CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A072001	Electromagnetic Interference		
							F26	No Health Consequences or Impact	A072001	Electromagnetic Interference		
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A072001	Electromagnetic Interference		
							F26	No Health Consequences or Impact	A072001	Electromagnetic Interference		
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A072001	Electromagnetic Interference		
							F26	No Health Consequences or Impact	A072001	Electromagnetic Interference		
17599	2025	All other reportable incidents	47265	Dual-chamber implantable pacemaker, rate-responsive	CT1685	Cardiac pacemakers	E0112	Dizziness	A0712	Pacing Problem		
							F11	Non-Serious Injury/ Illness/ Impairment	A0712	Pacing Problem		
					CT257	Implantable cardiac pacemakers	E0112	Dizziness	A0712	Pacing Problem		
							F11	Non-Serious Injury/ Illness/ Impairment	A0712	Pacing Problem		
					CT2866	Implantable cardiac electrical stimulators	E0112	Dizziness	A0712	Pacing Problem		
							F11	Non-Serious Injury/ Illness/ Impairment	A0712	Pacing Problem		
17600	2025	All other reportable incidents	47265	Dual-chamber implantable pacemaker, rate-responsive			E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
17601	2025	All other reportable incidents	47265	Dual-chamber implantable pacemaker, rate-responsive			E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
17602	2025	All other reportable incidents	60789	Intracardiac pacemaker			E2403	No Clinical Signs, Symptoms or Conditions	A070101	Failure to Capture		
									A1502	Positioning Problem		
							F26	No Health Consequences or Impact	A070101	Failure to Capture		
									A1502	Positioning Problem		
17603	2025	All other reportable incidents	47265	Dual-chamber implantable pacemaker, rate-responsive	CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A072001	Electromagnetic Interference		
							F26	No Health Consequences or Impact	A072001	Electromagnetic Interference		
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A072001	Electromagnetic Interference		
							F26	No Health Consequences or Impact	A072001	Electromagnetic Interference		
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A072001	Electromagnetic Interference		
							F26	No Health Consequences or Impact	A072001	Electromagnetic Interference		
17604	2025		47265	Dual-chamber implantable pacemaker, rate-responsive	CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A070501	Battery Problem: High Impedance		
							F0102	Therapeutic Response Increased	A070501	Battery Problem: High Impedance		
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A070501	Battery Problem: High Impedance		
							F0102	Therapeutic Response Increased	A070501	Battery Problem: High Impedance		

					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A070501	Battery Problem: High Impedance		
							F0102	Therapeutic Response Increased	A070501	Battery Problem: High Impedance		
17605	2025	All other reportable incidents	47265	Dual-chamber implantable pacemaker, rate-responsive			E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	10/11/2025	10/11/2025
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	10/11/2025	10/11/2025
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	10/11/2025	10/11/2025
17606	2025	All other reportable incidents	47265	Dual-chamber implantable pacemaker, rate-responsive			E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
17607	2025	All other reportable incidents			CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A072201	High Impedance		
							F2203	Imaging Required	A072201	High Impedance		
							F26	No Health Consequences or Impact	A072201	High Impedance		
17608	2025	All other reportable incidents	47265	Dual-chamber implantable pacemaker, rate-responsive			E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
									A0908	Incorrect, Inadequate or Imprecise Result or Readings		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
									A0908	Incorrect, Inadequate or Imprecise Result or Readings		
17609	2025	All other reportable incidents	47265	Dual-chamber implantable pacemaker, rate-responsive			E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
17610	2025		47265	Dual-chamber implantable pacemaker, rate-responsive	CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A070501	Battery Problem: High Impedance		
									A2501	Prophylactic Removal Due to Corrective Action		
							F1905	Device Revision or Replacement	A070501	Battery Problem: High Impedance		
									A2501	Prophylactic Removal Due to Corrective Action		
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A070501	Battery Problem: High Impedance		
									A2501	Prophylactic Removal Due to Corrective Action		
							F1905	Device Revision or Replacement	A070501	Battery Problem: High Impedance		
									A2501	Prophylactic Removal Due to Corrective Action		
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A070501	Battery Problem: High Impedance		
									A2501	Prophylactic Removal Due to Corrective Action		
							F1905	Device Revision or Replacement	A070501	Battery Problem: High Impedance		
									A2501	Prophylactic Removal Due to Corrective Action		
17611	2025	All other reportable incidents	47265	Dual-chamber implantable pacemaker, rate-responsive			E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
17612	2025	All other reportable incidents	47267	Single-chamber implantable pacemaker, rate-responsive			E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
17613	2025	All other reportable incidents	47267	Single-chamber implantable pacemaker, rate-responsive			E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
17614	2025	All other reportable incidents	47265	Dual-chamber implantable pacemaker, rate-responsive			E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
17615	2025		47265	Dual-chamber implantable pacemaker, rate-responsive	CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A0909	Unexpected Therapeutic Results		
									A27	Appropriate Device Problem Term/Code not Available		
							F23	Unexpected Medical Intervention	A0909	Unexpected Therapeutic Results		
									A27	Appropriate Device Problem Term/Code not Available		
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A0909	Unexpected Therapeutic Results		
									A27	Appropriate Device Problem Term/Code not Available		
							F23	Unexpected Medical Intervention	A0909	Unexpected Therapeutic Results		
									A27	Appropriate Device Problem Term/Code not Available		
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A0909	Unexpected Therapeutic Results		
									A27	Appropriate Device Problem Term/Code not Available		
							F23	Unexpected Medical Intervention	A0909	Unexpected Therapeutic Results		
									A27	Appropriate Device Problem Term/Code not Available		
17616	2025		47267	Single-chamber implantable pacemaker, rate-responsive	CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A27	Appropriate Device Problem Term/Code not Available		
							F0103	Unexpected Therapeutic Effects	A27	Appropriate Device Problem Term/Code not Available		
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A27	Appropriate Device Problem Term/Code not Available		
							F0103	Unexpected Therapeutic Effects	A27	Appropriate Device Problem Term/Code not Available		
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A27	Appropriate Device Problem Term/Code not Available		
							F0103	Unexpected Therapeutic Effects	A27	Appropriate Device Problem Term/Code not Available		
17617	2025		47265	Dual-chamber implantable pacemaker, rate-responsive	CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A26	Insufficient Device Problem Information		
							F0103	Unexpected Therapeutic Effects	A26	Insufficient Device Problem Information		
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A26	Insufficient Device Problem Information		
							F0103	Unexpected Therapeutic Effects	A26	Insufficient Device Problem Information		
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A26	Insufficient Device Problem Information		
							F0103	Unexpected Therapeutic Effects	A26	Insufficient Device Problem Information		
17618	2025	All other reportable incidents	47265	Dual-chamber implantable pacemaker, rate-responsive	CT1685	Cardiac pacemakers	E012207	Twitching	A0908	Incorrect, Inadequate or Imprecise Result or Readings	03/10/2025	03/10/2025
							E020201	Anxiety	A0908	Incorrect, Inadequate or Imprecise Result or Readings	03/10/2025	03/10/2025
							F08	Hospitalization or Prolonged Hospitalization	A0908	Incorrect, Inadequate or Imprecise Result or Readings	03/10/2025	03/10/2025
							F17	Sedation	A0908	Incorrect, Inadequate or Imprecise Result or Readings	03/10/2025	03/10/2025
							F1905	Device Revision or Replacement	A0908	Incorrect, Inadequate or Imprecise Result or Readings	03/10/2025	03/10/2025
					CT257	Implantable cardiac pacemakers	E012207	Twitching	A0908	Incorrect, Inadequate or Imprecise Result or Readings	03/10/2025	03/10/2025
							E020201	Anxiety	A0908	Incorrect, Inadequate or Imprecise Result or Readings	03/10/2025	03/10/2025
							F08	Hospitalization or Prolonged Hospitalization	A0908	Incorrect, Inadequate or Imprecise Result or Readings	03/10/2025	03/10/2025
							F17	Sedation	A0908	Incorrect, Inadequate or Imprecise Result or Readings	03/10/2025	03/10/2025

							F1905	Device Revision or Replacement	A0908	Incorrect, Inadequate or Imprecise Result or Readings	03/10/2025	03/10/2025
					CT2866	Implantable cardiac electrical stimulators	E012207	Twitching	A0908	Incorrect, Inadequate or Imprecise Result or Readings	03/10/2025	03/10/2025
							E020201	Anxiety	A0908	Incorrect, Inadequate or Imprecise Result or Readings	03/10/2025	03/10/2025
							F08	Hospitalization or Prolonged Hospitalization	A0908	Incorrect, Inadequate or Imprecise Result or Readings	03/10/2025	03/10/2025
							F17	Sedation	A0908	Incorrect, Inadequate or Imprecise Result or Readings	03/10/2025	03/10/2025
							F1905	Device Revision or Replacement	A0908	Incorrect, Inadequate or Imprecise Result or Readings	03/10/2025	03/10/2025
17619	2025	Death	63188	Left-ventricular wireless pacing system sensor/transmitter			E0505	Hematoma	A24	No Apparent Device Problem		
							F02	Death	A24	No Apparent Device Problem		
17620	2025	All other reportable incidents	47267	Single-chamber implantable pacemaker, rate-responsive			E2403	No Clinical Signs, Symptoms or Conditions	A05	Mechanical Problem	23/10/2025	23/10/2025
							F26	No Health Consequences or Impact	A05	Mechanical Problem	23/10/2025	23/10/2025
17621	2025	All other reportable incidents	47265	Dual-chamber implantable pacemaker, rate-responsive			E0601	Arrhythmia	A070910	Undersensing		
									A0712	Pacing Problem		
							F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A070910	Undersensing		
									A0712	Pacing Problem		
							F15	Recognized Device or Procedural Complication	A070910	Undersensing		
									A0712	Pacing Problem		
17622	2025	All other reportable incidents	47263	Cardiac resynchronization therapy implantable pacemaker	CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	01/10/2025	01/10/2025
									A0908	Incorrect, Inadequate or Imprecise Result or Readings	01/10/2025	01/10/2025
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	01/10/2025	01/10/2025
									A0908	Incorrect, Inadequate or Imprecise Result or Readings	01/10/2025	01/10/2025
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	01/10/2025	01/10/2025
									A0908	Incorrect, Inadequate or Imprecise Result or Readings	01/10/2025	01/10/2025
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	01/10/2025	01/10/2025
									A0908	Incorrect, Inadequate or Imprecise Result or Readings	01/10/2025	01/10/2025
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	01/10/2025	01/10/2025
									A0908	Incorrect, Inadequate or Imprecise Result or Readings	01/10/2025	01/10/2025
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	01/10/2025	01/10/2025
									A0908	Incorrect, Inadequate or Imprecise Result or Readings	01/10/2025	01/10/2025
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	01/10/2025	01/10/2025
									A0908	Incorrect, Inadequate or Imprecise Result or Readings	01/10/2025	01/10/2025
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	01/10/2025	01/10/2025
									A0908	Incorrect, Inadequate or Imprecise Result or Readings	01/10/2025	01/10/2025
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	01/10/2025	01/10/2025
									A0908	Incorrect, Inadequate or Imprecise Result or Readings	01/10/2025	01/10/2025
17623	2025		47263	Cardiac resynchronization therapy implantable pacemaker	CT1685	Cardiac pacemakers	E012207	Twitching	A071205	Pacemaker Found in Back-Up Mode		
									A0723	Unexpected Mode Switch		
							F2307	Reprogramming of Device	A071205	Pacemaker Found in Back-Up Mode		
									A0723	Unexpected Mode Switch		
					CT257	Implantable cardiac pacemakers	E012207	Twitching	A071205	Pacemaker Found in Back-Up Mode		
									A0723	Unexpected Mode Switch		
							F2307	Reprogramming of Device	A071205	Pacemaker Found in Back-Up Mode		
									A0723	Unexpected Mode Switch		
					CT2866	Implantable cardiac electrical stimulators	E012207	Twitching	A071205	Pacemaker Found in Back-Up Mode		
									A0723	Unexpected Mode Switch		
							F2307	Reprogramming of Device	A071205	Pacemaker Found in Back-Up Mode		
									A0723	Unexpected Mode Switch		
17624	2025	All other reportable incidents	47265	Dual-chamber implantable pacemaker, rate-responsive			E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
17625	2025	All other reportable incidents	37265	Dual-chamber implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A160602	Premature End-of-Life Indicator	06/10/2025	06/10/2025
							F1905	Device Revision or Replacement	A160602	Premature End-of-Life Indicator	06/10/2025	06/10/2025
17626	2025	All other reportable incidents	47265	Dual-chamber implantable pacemaker, rate-responsive			E0601	Arrhythmia	A0712	Pacing Problem		
							E060109	Tachycardia	A0712	Pacing Problem		
							F12	Serious Injury/ Illness/ Impairment	A0712	Pacing Problem		
							F2307	Reprogramming of Device	A0712	Pacing Problem		
17627	2025		47263	Cardiac resynchronization therapy implantable pacemaker	CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem		
							F19	Surgical Intervention	A0705	Battery Problem		
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem		
							F19	Surgical Intervention	A0705	Battery Problem		
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem		
							F19	Surgical Intervention	A0705	Battery Problem		
17628	2025	All other reportable incidents	47265	Dual-chamber implantable pacemaker, rate-responsive			E2403	No Clinical Signs, Symptoms or Conditions	A070101	Failure to Capture		
							F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A070101	Failure to Capture		
17629	2025	All other reportable incidents	37265	Dual-chamber implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A070910	Undersensing		
									A071302	Inappropriate/Inadequate Shock/Stimulation		
							F12	Serious Injury/ Illness/ Impairment	A070910	Undersensing		
									A071302	Inappropriate/Inadequate Shock/Stimulation		
17630	2025	All other reportable incidents	47265	Dual-chamber implantable pacemaker, rate-responsive			E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
17631	2025	All other reportable incidents	47265	Dual-chamber implantable pacemaker, rate-responsive			E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
17632	2025	All other reportable incidents	47265	Dual-chamber implantable pacemaker, rate-responsive			E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
17633	2025	Death	17882	Physiologic-monitoring defibrillation system	CT1684	Cardiac pacemakers and associated devices	E2401	Insufficient Information	A26	Insufficient Device Problem Information		
							F05	Delay to Treatment/ Therapy	A26	Insufficient Device Problem Information		
							F2301	Additional Device Required	A26	Insufficient Device Problem Information		
17634	2025	All other reportable incidents	47267	Single-chamber implantable pacemaker, rate-			E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		

				responsive			F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
17635	2025	All other reportable incidents	47265	Dual-chamber implantable pacemaker, rate-responsive			E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
17636	2025	All other reportable incidents	47265	Dual-chamber implantable pacemaker, rate-responsive			E2103	Device Overstimulation of Tissue	A071209	Extracardiac Stimulation		
							F12	Serious Injury/ Illness/ Impairment	A071209	Extracardiac Stimulation		
							F2307	Reprogramming of Device	A071209	Extracardiac Stimulation		
17637	2025	All other reportable incidents	47267	Single-chamber implantable pacemaker, rate-responsive			E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem		
							F26	No Health Consequences or Impact	A0705	Battery Problem		
17638	2025		47263	Cardiac resynchronization therapy implantable pacemaker	CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A0203	Defective Device		
									A070501	Battery Problem: High Impedance		
							F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A0203	Defective Device		
									A070501	Battery Problem: High Impedance		
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A0203	Defective Device		
									A070501	Battery Problem: High Impedance		
							F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A0203	Defective Device		
									A070501	Battery Problem: High Impedance		
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A0203	Defective Device		
									A070501	Battery Problem: High Impedance		
							F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A0203	Defective Device		
									A070501	Battery Problem: High Impedance		
17639	2025	All other reportable incidents	47265	Dual-chamber implantable pacemaker, rate-responsive			E2403	No Clinical Signs, Symptoms or Conditions	A070909	Oversensing		
									A072201	High Impedance		
							F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A070909	Oversensing		
									A072201	High Impedance		
17640	2025	All other reportable incidents	60789	Intracardiac pacemaker			E2402	Appropriate Clinical Signs, Symptoms and Conditions Term/Code not Available	A072201	High Impedance		
							F28	Appropriate Health Impact Term/Code not Available	A072201	High Impedance		
17641	2025	All other reportable incidents	47265	Dual-chamber implantable pacemaker, rate-responsive			E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	01/12/2025	01/12/2025
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	01/12/2025	01/12/2025
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	01/12/2025	01/12/2025
17642	2025	All other reportable incidents	47267	Single-chamber implantable pacemaker, rate-responsive			E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
17643	2025	All other reportable incidents	47267	Single-chamber implantable pacemaker, rate-responsive			E2403	No Clinical Signs, Symptoms or Conditions	A0908	Incorrect, Inadequate or Imprecise Result or Readings	30/10/2025	30/10/2025
							F08	Hospitalization or Prolonged Hospitalization	A0908	Incorrect, Inadequate or Imprecise Result or Readings	30/10/2025	30/10/2025
							F1905	Device Revision or Replacement	A0908	Incorrect, Inadequate or Imprecise Result or Readings	30/10/2025	30/10/2025
17644	2025		47265	Dual-chamber implantable pacemaker, rate-responsive	CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	09/09/2025	09/09/2025
							F1901	Additional Surgery	A24	No Apparent Device Problem	09/09/2025	09/09/2025
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	09/09/2025	09/09/2025
							F1901	Additional Surgery	A24	No Apparent Device Problem	09/09/2025	09/09/2025
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	09/09/2025	09/09/2025
							F1901	Additional Surgery	A24	No Apparent Device Problem	09/09/2025	09/09/2025
17645	2025	All other reportable incidents	47267	Single-chamber implantable pacemaker, rate-responsive			E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
17646	2025	All other reportable incidents	47265	Dual-chamber implantable pacemaker, rate-responsive			E2403	No Clinical Signs, Symptoms or Conditions	A072001	Electromagnetic Interference		
							F26	No Health Consequences or Impact	A072001	Electromagnetic Interference		
17647	2025	Unanticipated serious deterioration in state of health	47265	Dual-chamber implantable pacemaker, rate-responsive			E0601	Arrhythmia	A0712	Pacing Problem		
							E0618	Non Specific EKG/ECG Changes	A0712	Pacing Problem		
							F12	Serious Injury/ Illness/ Impairment	A0712	Pacing Problem		
							F2307	Reprogramming of Device	A0712	Pacing Problem		
							F26	No Health Consequences or Impact	A0712	Pacing Problem		
17648	2025		47265	Dual-chamber implantable pacemaker, rate-responsive	CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A160503	Inappropriate or Unexpected Reset		
							F1901	Additional Surgery	A160503	Inappropriate or Unexpected Reset		
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A160503	Inappropriate or Unexpected Reset		
							F1901	Additional Surgery	A160503	Inappropriate or Unexpected Reset		
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A160503	Inappropriate or Unexpected Reset		
							F1901	Additional Surgery	A160503	Inappropriate or Unexpected Reset		
17649	2025		47263	Cardiac resynchronization therapy implantable pacemaker	CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F1901	Additional Surgery	A070504	Premature Discharge of Battery		
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F1901	Additional Surgery	A070504	Premature Discharge of Battery		
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F1901	Additional Surgery	A070504	Premature Discharge of Battery		
17650	2025		47263	Cardiac resynchronization therapy implantable pacemaker	CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A02	Manufacturing, Packaging or Shipping Problem		
							F04	Delay to Diagnosis	A02	Manufacturing, Packaging or Shipping Problem		
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A02	Manufacturing, Packaging or Shipping Problem		
							F04	Delay to Diagnosis	A02	Manufacturing, Packaging or Shipping Problem		
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A02	Manufacturing, Packaging or Shipping Problem		
							F04	Delay to Diagnosis	A02	Manufacturing, Packaging or Shipping Problem		
17651	2025	All other reportable incidents	47265	Dual-chamber implantable pacemaker, rate-responsive			E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
17652	2025	All other reportable incidents	47265	Dual-chamber implantable pacemaker, rate-responsive			E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
17653	2025	All other reportable incidents	47265	Dual-chamber implantable pacemaker, rate-responsive			E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem		
									A0908	Incorrect, Inadequate or Imprecise Result or Readings		
									A1107	Data Problem		
									A1304	Telemetry Discrepancy		
							F26	No Health Consequences or Impact	A0705	Battery Problem		
									A0908	Incorrect, Inadequate or Imprecise Result or Readings		
									A1107	Data Problem		
									A1304	Telemetry Discrepancy		
17654	2025	All other reportable incidents	47265	Dual-chamber implantable pacemaker, rate-responsive			E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
17655	2025	All other reportable incidents	47265	Dual-chamber implantable pacemaker, rate-responsive			E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		

17656	2025	All other reportable incidents	47265	Dual-chamber implantable pacemaker, rate-responsive			E2403	No Clinical Signs, Symptoms or Conditions	A072201	High Impedance	17/10/2025	17/10/2025
									A12	Connection Problem	17/10/2025	17/10/2025
							F08	Hospitalization or Prolonged Hospitalization	A072201	High Impedance	17/10/2025	17/10/2025
									A12	Connection Problem	17/10/2025	17/10/2025
							F1905	Device Revision or Replacement	A072201	High Impedance	17/10/2025	17/10/2025
									A12	Connection Problem	17/10/2025	17/10/2025
17657	2025	All other reportable incidents	47265	Dual-chamber implantable pacemaker, rate-responsive			E2403	No Clinical Signs, Symptoms or Conditions	A070909	Oversensing		
									A072201	High Impedance		
							F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A070909	Oversensing		
									A072201	High Impedance		
							F2203	Imaging Required	A070909	Oversensing		
									A072201	High Impedance		
17658	2025	All other reportable incidents	47265	Dual-chamber implantable pacemaker, rate-responsive			E2403	No Clinical Signs, Symptoms or Conditions	A05	Mechanical Problem		
							F26	No Health Consequences or Impact	A05	Mechanical Problem		
17659	2025	All other reportable incidents	47267	Single-chamber implantable pacemaker, rate-responsive			E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
17660	2025	All other reportable incidents	47265	Dual-chamber implantable pacemaker, rate-responsive			E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	19/12/2025	19/12/2025
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	19/12/2025	19/12/2025
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	19/12/2025	19/12/2025
17661	2025	Death	17882	Physiologic-monitoring defibrillation system	CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070903	Failure to Analyze Signal		
							F02	Death	A070903	Failure to Analyze Signal		
17662	2025	All other reportable incidents	47270	Cardiac resynchronization therapy implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E0601	Arrhythmia	A070504	Premature Discharge of Battery	10/11/2025	10/11/2025
									A071102	Failure to Interrogate	10/11/2025	10/11/2025
							F1901	Additional Surgery	A070504	Premature Discharge of Battery	10/11/2025	10/11/2025
									A071102	Failure to Interrogate	10/11/2025	10/11/2025
17663	2025		17882	Physiologic-monitoring defibrillation system	CT1684	Cardiac pacemakers and associated devices	E0602	Cardiac Arrest	A1301	Failure to Read Input Signal		
							F02	Death	A1301	Failure to Read Input Signal		
17664	2025	All other reportable incidents	47265	Dual-chamber implantable pacemaker, rate-responsive	CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	14/08/2025	14/08/2025
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	14/08/2025	14/08/2025
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	14/08/2025	14/08/2025
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	14/08/2025	14/08/2025
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	14/08/2025	14/08/2025
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	14/08/2025	14/08/2025
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	14/08/2025	14/08/2025
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	14/08/2025	14/08/2025
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	14/08/2025	14/08/2025
17665	2025		47265	Dual-chamber implantable pacemaker, rate-responsive	CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A11	Computer Software Problem		
									A110207	Problem with Software Installation		
							F26	No Health Consequences or Impact	A11	Computer Software Problem		
									A110207	Problem with Software Installation		
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A11	Computer Software Problem		
									A110207	Problem with Software Installation		
							F26	No Health Consequences or Impact	A11	Computer Software Problem		
									A110207	Problem with Software Installation		
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A11	Computer Software Problem		
									A110207	Problem with Software Installation		
							F26	No Health Consequences or Impact	A11	Computer Software Problem		
									A110207	Problem with Software Installation		
17666	2025	All other reportable incidents	47267	Single-chamber implantable pacemaker, rate-responsive			E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
17667	2025	All other reportable incidents	47265	Dual-chamber implantable pacemaker, rate-responsive			E2403	No Clinical Signs, Symptoms or Conditions	A0908	Incorrect, Inadequate or Imprecise Result or Readings	23/10/2025	23/10/2025
							F08	Hospitalization or Prolonged Hospitalization	A0908	Incorrect, Inadequate or Imprecise Result or Readings	23/10/2025	23/10/2025
							F1905	Device Revision or Replacement	A0908	Incorrect, Inadequate or Imprecise Result or Readings	23/10/2025	23/10/2025
17668	2025	All other reportable incidents	47265	Dual-chamber implantable pacemaker, rate-responsive			E2403	No Clinical Signs, Symptoms or Conditions	A070909	Oversensing		
							F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A070909	Oversensing		
17669	2025	All other reportable incidents	60789	Intracardiac pacemaker			E2403	No Clinical Signs, Symptoms or Conditions	A09	Output Problem	07/11/2025	07/11/2025
									A1107	Data Problem	07/11/2025	07/11/2025
							F26	No Health Consequences or Impact	A09	Output Problem	07/11/2025	07/11/2025
									A1107	Data Problem	07/11/2025	07/11/2025
17670	2025	All other reportable incidents	47265	Dual-chamber implantable pacemaker, rate-responsive	CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
17671	2025	All other reportable incidents	47265	Dual-chamber implantable pacemaker, rate-responsive			E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	25/11/2025	25/11/2025
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery	25/11/2025	25/11/2025
17672	2025		47265	Dual-chamber implantable pacemaker, rate-responsive	CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A071205	Pacemaker Found in Back-Up Mode		
							F1905	Device Revision or Replacement	A071205	Pacemaker Found in Back-Up Mode		
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A071205	Pacemaker Found in Back-Up Mode		
							F1905	Device Revision or Replacement	A071205	Pacemaker Found in Back-Up Mode		
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A071205	Pacemaker Found in Back-Up Mode		
							F1905	Device Revision or Replacement	A071205	Pacemaker Found in Back-Up Mode		
17673	2025	All other reportable incidents	47267	Single-chamber implantable pacemaker, rate-responsive			E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
17674	2025	All other reportable incidents	47265	Dual-chamber implantable pacemaker, rate-responsive			E2403	No Clinical Signs, Symptoms or Conditions	A071102	Failure to Interrogate	10/11/2025	10/11/2025
							F26	No Health Consequences or Impact	A071102	Failure to Interrogate	10/11/2025	10/11/2025
17675	2025	All other reportable incidents	47270	Cardiac resynchronization therapy implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E011903	Syncope/Fainting	A071201	Failure to Convert Rhythm		
							E0601	Arrhythmia	A071201	Failure to Convert Rhythm		
							F12	Serious Injury/ Illness/ Impairment	A071201	Failure to Convert Rhythm		
17676	2025		47265	Dual-chamber implantable pacemaker, rate-responsive	CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions				
									A0501	Detachment of Device or Device Component		
							F26	No Health Consequences or Impact				
									A0501	Detachment of Device or Device Component		
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions				

									A0501	Detachment of Device or Device Component		
							F26	No Health Consequences or Impact				
									A0501	Detachment of Device or Device Component		
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions				
									A0501	Detachment of Device or Device Component		
							F26	No Health Consequences or Impact				
									A0501	Detachment of Device or Device Component		
17677	2025	All other reportable incidents	47265	Dual-chamber implantable pacemaker, rate-responsive			E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	24/11/2025	24/11/2025
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	24/11/2025	24/11/2025
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	24/11/2025	24/11/2025
17678	2025	All other reportable incidents	47265	Dual-chamber implantable pacemaker, rate-responsive			E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
17679	2025		47265	Dual-chamber implantable pacemaker, rate-responsive	CT1685	Cardiac pacemakers	E012207	Twitching	A071205	Pacemaker Found in Back-Up Mode		
									A0909	Unexpected Therapeutic Results		
									A160503	Inappropriate or Unexpected Reset		
							F01	Change in Therapeutic Response	A071205	Pacemaker Found in Back-Up Mode		
									A0909	Unexpected Therapeutic Results		
									A160503	Inappropriate or Unexpected Reset		
							F25	Unanticipated Adverse Device Effect	A071205	Pacemaker Found in Back-Up Mode		
									A0909	Unexpected Therapeutic Results		
									A160503	Inappropriate or Unexpected Reset		
					CT257	Implantable cardiac pacemakers	E012207	Twitching	A071205	Pacemaker Found in Back-Up Mode		
									A0909	Unexpected Therapeutic Results		
									A160503	Inappropriate or Unexpected Reset		
							F01	Change in Therapeutic Response	A071205	Pacemaker Found in Back-Up Mode		
									A0909	Unexpected Therapeutic Results		
									A160503	Inappropriate or Unexpected Reset		
							F25	Unanticipated Adverse Device Effect	A071205	Pacemaker Found in Back-Up Mode		
									A0909	Unexpected Therapeutic Results		
									A160503	Inappropriate or Unexpected Reset		
					CT2866	Implantable cardiac electrical stimulators	E012207	Twitching	A071205	Pacemaker Found in Back-Up Mode		
									A0909	Unexpected Therapeutic Results		
									A160503	Inappropriate or Unexpected Reset		
							F01	Change in Therapeutic Response	A071205	Pacemaker Found in Back-Up Mode		
									A0909	Unexpected Therapeutic Results		
									A160503	Inappropriate or Unexpected Reset		
							F25	Unanticipated Adverse Device Effect	A071205	Pacemaker Found in Back-Up Mode		
									A0909	Unexpected Therapeutic Results		
									A160503	Inappropriate or Unexpected Reset		
17680	2025	All other reportable incidents	47265	Dual-chamber implantable pacemaker, rate-responsive			E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
17681	2025		37265	Dual-chamber implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem		
							F1905	Device Revision or Replacement	A0705	Battery Problem		
17682	2025		17882	Physiologic-monitoring defibrillation system	CT1684	Cardiac pacemakers and associated devices	E2401	Insufficient Information	A0705	Battery Problem		
									A071301	Failure to Deliver Shock/Stimulation		
							F05	Delay to Treatment/ Therapy	A0705	Battery Problem		
									A071301	Failure to Deliver Shock/Stimulation		
17683	2025	All other reportable incidents	60789	Intracardiac pacemaker			E2403	No Clinical Signs, Symptoms or Conditions	A1107	Data Problem		
							F26	No Health Consequences or Impact	A1107	Data Problem		
17684	2025		47265	Dual-chamber implantable pacemaker, rate-responsive	CT1685	Cardiac pacemakers	E011903	Syncope/Fainting	A0203	Defective Device		
									A070101	Failure to Capture		
									A070504	Premature Discharge of Battery		
									A071204	No Pacing		
							F08	Hospitalization or Prolonged Hospitalization	A0203	Defective Device		
									A070101	Failure to Capture		
									A070504	Premature Discharge of Battery		
									A071204	No Pacing		
					CT257	Implantable cardiac pacemakers	E011903	Syncope/Fainting	A0203	Defective Device		
									A070101	Failure to Capture		
									A070504	Premature Discharge of Battery		
									A071204	No Pacing		
							F08	Hospitalization or Prolonged Hospitalization	A0203	Defective Device		
									A070101	Failure to Capture		
									A070504	Premature Discharge of Battery		
									A071204	No Pacing		
					CT2866	Implantable cardiac electrical stimulators	E011903	Syncope/Fainting	A0203	Defective Device		
									A070101	Failure to Capture		
									A070504	Premature Discharge of Battery		
									A071204	No Pacing		
							F08	Hospitalization or Prolonged Hospitalization	A0203	Defective Device		
									A070101	Failure to Capture		
									A070504	Premature Discharge of Battery		
									A071204	No Pacing		
17685	2025	All other reportable incidents	47265	Dual-chamber implantable pacemaker, rate-responsive			E2311	Discomfort	A070909	Oversensing		
							F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A070909	Oversensing		
17686	2025	All other reportable incidents	47265	Dual-chamber implantable pacemaker, rate-responsive			E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
17687	2025	All other reportable incidents	60789	Intracardiac pacemaker			E0112	Dizziness	A070101	Failure to Capture		
									A071102	Failure to Interrogate		
									A12	Connection Problem		
							E011903	Syncope/Fainting	A070101	Failure to Capture		
									A071102	Failure to Interrogate		
									A12	Connection Problem		
							E0717	Dyspnea	A070101	Failure to Capture		
									A071102	Failure to Interrogate		
									A12	Connection Problem		
							F1901	Additional Surgery	A070101	Failure to Capture		
									A071102	Failure to Interrogate		

									A12	Connection Problem		
17688	2025	All other reportable incidents	47265	Dual-chamber implantable pacemaker, rate-responsive	CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
									A071205	Pacemaker Found in Back-Up Mode		
									A160503	Inappropriate or Unexpected Reset		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
									A071205	Pacemaker Found in Back-Up Mode		
									A160503	Inappropriate or Unexpected Reset		
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
									A071205	Pacemaker Found in Back-Up Mode		
									A160503	Inappropriate or Unexpected Reset		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
									A071205	Pacemaker Found in Back-Up Mode		
									A160503	Inappropriate or Unexpected Reset		
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
									A071205	Pacemaker Found in Back-Up Mode		
									A160503	Inappropriate or Unexpected Reset		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
									A071205	Pacemaker Found in Back-Up Mode		
									A160503	Inappropriate or Unexpected Reset		
17689	2025	All other reportable incidents	60789	Intracardiac pacemaker			E2403	No Clinical Signs, Symptoms or Conditions	A070901	Decreased Sensitivity	13/11/2025	13/11/2025
							F26	No Health Consequences or Impact	A070901	Decreased Sensitivity	13/11/2025	13/11/2025
17690	2025	All other reportable incidents	47265	Dual-chamber implantable pacemaker, rate-responsive			E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
17691	2025	All other reportable incidents	47265	Dual-chamber implantable pacemaker, rate-responsive	CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A020102	Nonstandard Device	27/03/2025	27/03/2025
							F24	Insufficient Health Impact Information	A020102	Nonstandard Device	27/03/2025	27/03/2025
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A020102	Nonstandard Device	27/03/2025	27/03/2025
							F24	Insufficient Health Impact Information	A020102	Nonstandard Device	27/03/2025	27/03/2025
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A020102	Nonstandard Device	27/03/2025	27/03/2025
							F24	Insufficient Health Impact Information	A020102	Nonstandard Device	27/03/2025	27/03/2025
17692	2025		17882	Physiologic-monitoring defibrillation system	CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A110401	Operating System Becomes Nonfunctional		
									A150105	Key or Button Unresponsive/not Working		
							F2601	Problem Identified Before Clinical Use/Exposure	A110401	Operating System Becomes Nonfunctional		
									A150105	Key or Button Unresponsive/not Working		
17693	2025	All other reportable incidents	60789	Intracardiac pacemaker			E1020	Nausea	A0712	Pacing Problem		
							E233001	Chest Pain	A0712	Pacing Problem		
							F12	Serious Injury/ Illness/ Impairment	A0712	Pacing Problem		
							F2307	Reprogramming of Device	A0712	Pacing Problem		
17694	2025	All other reportable incidents	47267	Single-chamber implantable pacemaker, rate-responsive			E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
17695	2025	All other reportable incidents	47267	Single-chamber implantable pacemaker, rate-responsive			E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem		
							F26	No Health Consequences or Impact	A0705	Battery Problem		
17696	2025	All other reportable incidents	47265	Dual-chamber implantable pacemaker, rate-responsive			E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
17697	2025	All other reportable incidents	60789	Intracardiac pacemaker			E2403	No Clinical Signs, Symptoms or Conditions	A1107	Data Problem		
							F26	No Health Consequences or Impact	A1107	Data Problem		
17698	2025		47265	Dual-chamber implantable pacemaker, rate-responsive	CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem		
							F26	No Health Consequences or Impact	A0705	Battery Problem		
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem		
							F26	No Health Consequences or Impact	A0705	Battery Problem		
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem		
							F26	No Health Consequences or Impact	A0705	Battery Problem		
17699	2025	All other reportable incidents	47265	Dual-chamber implantable pacemaker, rate-responsive			E2403	No Clinical Signs, Symptoms or Conditions	A020102	Nonstandard Device	23/10/2025	23/10/2025
									A070504	Premature Discharge of Battery	23/10/2025	23/10/2025
							F08	Hospitalization or Prolonged Hospitalization	A020102	Nonstandard Device	23/10/2025	23/10/2025
									A070504	Premature Discharge of Battery	23/10/2025	23/10/2025
							F1905	Device Revision or Replacement	A020102	Nonstandard Device	23/10/2025	23/10/2025
									A070504	Premature Discharge of Battery	23/10/2025	23/10/2025
17700	2025	All other reportable incidents	47267	Single-chamber implantable pacemaker, rate-responsive			E020204	Malaise	A26	Insufficient Device Problem Information		
							F26	No Health Consequences or Impact	A26	Insufficient Device Problem Information		
17701	2025	All other reportable incidents	63189	Implantable pacemaker battery	CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
17702	2025	All other reportable incidents	47265	Dual-chamber implantable pacemaker, rate-responsive			E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	04/12/2025	04/12/2025
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	04/12/2025	04/12/2025
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	04/12/2025	04/12/2025
17703	2025	All other reportable incidents	60789	Intracardiac pacemaker			E2403	No Clinical Signs, Symptoms or Conditions	A1107	Data Problem	19/11/2025	19/11/2025
							F26	No Health Consequences or Impact	A1107	Data Problem	19/11/2025	19/11/2025
17704	2025	All other reportable incidents	47265	Dual-chamber implantable pacemaker, rate-responsive			E2403	No Clinical Signs, Symptoms or Conditions	A05	Mechanical Problem	19/11/2025	19/11/2025
							F08	Hospitalization or Prolonged Hospitalization	A05	Mechanical Problem	19/11/2025	19/11/2025
							F12	Serious Injury/ Illness/ Impairment	A05	Mechanical Problem	19/11/2025	19/11/2025
							F1905	Device Revision or Replacement	A05	Mechanical Problem	19/11/2025	19/11/2025
17705	2025	All other reportable incidents	47265	Dual-chamber implantable pacemaker, rate-responsive			E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
17706	2025		47263	Cardiac resynchronization therapy implantable pacemaker	CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A0203	Defective Device		
									A11	Computer Software Problem		
							F19	Surgical Intervention	A0203	Defective Device		
									A11	Computer Software Problem		
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A0203	Defective Device		
									A11	Computer Software Problem		
							F19	Surgical Intervention	A0203	Defective Device		
									A11	Computer Software Problem		
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A0203	Defective Device		
									A11	Computer Software Problem		
							F19	Surgical Intervention	A0203	Defective Device		
									A11	Computer Software Problem		

17707	2025	All other reportable incidents	47265	Dual-chamber implantable pacemaker, rate-responsive			E2403	No Clinical Signs, Symptoms or Conditions	A0908	Incorrect, Inadequate or Imprecise Result or Readings	05/11/2025	05/11/2025
							F08	Hospitalization or Prolonged Hospitalization	A0908	Incorrect, Inadequate or Imprecise Result or Readings	05/11/2025	05/11/2025
							F1905	Device Revision or Replacement	A0908	Incorrect, Inadequate or Imprecise Result or Readings	05/11/2025	05/11/2025
17708	2025	All other reportable incidents	47265	Dual-chamber implantable pacemaker, rate-responsive			E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
17709	2025	All other reportable incidents	47265	Dual-chamber implantable pacemaker, rate-responsive			E2403	No Clinical Signs, Symptoms or Conditions	A0908	Incorrect, Inadequate or Imprecise Result or Readings	01/01/2026	01/01/2026
							F08	Hospitalization or Prolonged Hospitalization	A0908	Incorrect, Inadequate or Imprecise Result or Readings	01/01/2026	01/01/2026
							F1905	Device Revision or Replacement	A0908	Incorrect, Inadequate or Imprecise Result or Readings	01/01/2026	01/01/2026
17710	2025	All other reportable incidents	47267	Single-chamber implantable pacemaker, rate-responsive			E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	01/01/2026	01/01/2026
									A072201	High Impedance	01/01/2026	01/01/2026
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	01/01/2026	01/01/2026
									A072201	High Impedance	01/01/2026	01/01/2026
							F1903	Device Explantation	A070504	Premature Discharge of Battery	01/01/2026	01/01/2026
									A072201	High Impedance	01/01/2026	01/01/2026
17711	2025		47265	Dual-chamber implantable pacemaker, rate-responsive			E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	13/09/2025	13/09/2025
							F1901	Additional Surgery	A24	No Apparent Device Problem	13/09/2025	13/09/2025
17712	2025	All other reportable incidents	60789	Intracardiac pacemaker			E2403	No Clinical Signs, Symptoms or Conditions	A051201	Device Dislodged or Dislocated		
									A150302	Difficult or Delayed Separation		
							F1908	Prolonged Surgery	A051201	Device Dislodged or Dislocated		
									A150302	Difficult or Delayed Separation		
17713	2025	All other reportable incidents	47265	Dual-chamber implantable pacemaker, rate-responsive			E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
17714	2025	All other reportable incidents	47265	Dual-chamber implantable pacemaker, rate-responsive	CT1685	Cardiac pacemakers	E2401	Insufficient Information	A070504	Premature Discharge of Battery		
									A071205	Pacemaker Found in Back-Up Mode		
							E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
									A071205	Pacemaker Found in Back-Up Mode		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
									A071205	Pacemaker Found in Back-Up Mode		
					CT257	Implantable cardiac pacemakers	E2401	Insufficient Information	A070504	Premature Discharge of Battery		
									A071205	Pacemaker Found in Back-Up Mode		
							E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
									A071205	Pacemaker Found in Back-Up Mode		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
									A071205	Pacemaker Found in Back-Up Mode		
					CT2866	Implantable cardiac electrical stimulators	E2401	Insufficient Information	A070504	Premature Discharge of Battery		
									A071205	Pacemaker Found in Back-Up Mode		
							E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
									A071205	Pacemaker Found in Back-Up Mode		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
									A071205	Pacemaker Found in Back-Up Mode		
17715	2025	All other reportable incidents	47265	Dual-chamber implantable pacemaker, rate-responsive			E2403	No Clinical Signs, Symptoms or Conditions	A05	Mechanical Problem	19/11/2025	19/11/2025
							F26	No Health Consequences or Impact	A05	Mechanical Problem	19/11/2025	19/11/2025
17716	2025	All other reportable incidents	47265	Dual-chamber implantable pacemaker, rate-responsive	CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A0908	Incorrect, Inadequate or Imprecise Result or Readings		
							F08	Hospitalization or Prolonged Hospitalization	A0908	Incorrect, Inadequate or Imprecise Result or Readings		
							F1905	Device Revision or Replacement	A0908	Incorrect, Inadequate or Imprecise Result or Readings		
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A0908	Incorrect, Inadequate or Imprecise Result or Readings		
							F08	Hospitalization or Prolonged Hospitalization	A0908	Incorrect, Inadequate or Imprecise Result or Readings		
							F1905	Device Revision or Replacement	A0908	Incorrect, Inadequate or Imprecise Result or Readings		
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A0908	Incorrect, Inadequate or Imprecise Result or Readings		
							F08	Hospitalization or Prolonged Hospitalization	A0908	Incorrect, Inadequate or Imprecise Result or Readings		
							F1905	Device Revision or Replacement	A0908	Incorrect, Inadequate or Imprecise Result or Readings		
17717	2025	All other reportable incidents	47265	Dual-chamber implantable pacemaker, rate-responsive			E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
17718	2022	All other reportable incidents	47263	Cardiac resynchronization therapy implantable pacemaker	CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A0908	Incorrect, Inadequate or Imprecise Result or Readings		
							F08	Hospitalization or Prolonged Hospitalization	A0908	Incorrect, Inadequate or Imprecise Result or Readings		
							F1905	Device Revision or Replacement	A0908	Incorrect, Inadequate or Imprecise Result or Readings		
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A0908	Incorrect, Inadequate or Imprecise Result or Readings		
							F08	Hospitalization or Prolonged Hospitalization	A0908	Incorrect, Inadequate or Imprecise Result or Readings		
							F1905	Device Revision or Replacement	A0908	Incorrect, Inadequate or Imprecise Result or Readings		
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A0908	Incorrect, Inadequate or Imprecise Result or Readings		
							F08	Hospitalization or Prolonged Hospitalization	A0908	Incorrect, Inadequate or Imprecise Result or Readings		
							F1905	Device Revision or Replacement	A0908	Incorrect, Inadequate or Imprecise Result or Readings		
17719	2023		47265	Dual-chamber implantable pacemaker, rate-responsive	CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	26/02/2023	26/02/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	26/02/2023	26/02/2023
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	26/02/2023	26/02/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	26/02/2023	26/02/2023
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	26/02/2023	26/02/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	26/02/2023	26/02/2023
17720	2023	All other reportable incidents	47265	Dual-chamber implantable pacemaker, rate-	CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	31/05/2023	31/05/2023

				responsive			F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	31/05/2023	31/05/2023
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	31/05/2023	31/05/2023
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	31/05/2023	31/05/2023
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	31/05/2023	31/05/2023
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	31/05/2023	31/05/2023
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	31/05/2023	31/05/2023
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	31/05/2023	31/05/2023
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	31/05/2023	31/05/2023
17721	2023	All other reportable incidents	35852	Single-chamber implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E2104	Electric Shock	A0709	Device Sensing Problem		
									A070909	Oversensing		
									A071302	Inappropriate/Inadequate Shock/Stimulation		
							F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A0709	Device Sensing Problem		
									A070909	Oversensing		
									A071302	Inappropriate/Inadequate Shock/Stimulation		
17722	2020	All other reportable incidents	47265	Dual-chamber implantable pacemaker, rate-responsive	CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
17723	2020	All other reportable incidents	47270	Cardiac resynchronization therapy implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A071302	Inappropriate/Inadequate Shock/Stimulation		
							F26	No Health Consequences or Impact	A071302	Inappropriate/Inadequate Shock/Stimulation		
17724	2020	All other reportable incidents	47270	Cardiac resynchronization therapy implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A070603	Failure to Charge	18/11/2020	18/11/2020
							F1901	Additional Surgery	A070603	Failure to Charge	18/11/2020	18/11/2020
17725	2023	All other reportable incidents	37265	Dual-chamber implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A07	Electrical /Electronic Property Problem	05/05/2023	05/05/2023
									A070504	Premature Discharge of Battery	05/05/2023	05/05/2023
							F08	Hospitalization or Prolonged Hospitalization	A07	Electrical /Electronic Property Problem	05/05/2023	05/05/2023
									A070504	Premature Discharge of Battery	05/05/2023	05/05/2023
							F12	Serious Injury/ Illness/ Impairment	A07	Electrical /Electronic Property Problem	05/05/2023	05/05/2023
									A070504	Premature Discharge of Battery	05/05/2023	05/05/2023
							F1905	Device Revision or Replacement	A07	Electrical /Electronic Property Problem	05/05/2023	05/05/2023
									A070504	Premature Discharge of Battery	05/05/2023	05/05/2023
17726	2024	Death	47263	Cardiac resynchronization therapy implantable pacemaker	CT1685	Cardiac pacemakers	E060109	Tachycardia	A24	No Apparent Device Problem	01/01/2024	31/12/2024
							F02	Death	A24	No Apparent Device Problem	01/01/2024	31/12/2024
					CT257	Implantable cardiac pacemakers	E060109	Tachycardia	A24	No Apparent Device Problem	01/01/2024	31/12/2024
							F02	Death	A24	No Apparent Device Problem	01/01/2024	31/12/2024
					CT2866	Implantable cardiac electrical stimulators	E060109	Tachycardia	A24	No Apparent Device Problem	01/01/2024	31/12/2024
							F02	Death	A24	No Apparent Device Problem	01/01/2024	31/12/2024
17727	2024	All other reportable incidents	35852	Single-chamber implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	23/04/2024	23/04/2024
									A0908	Incorrect, Inadequate or Imprecise Result or Readings	23/04/2024	23/04/2024
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	23/04/2024	23/04/2024
									A0908	Incorrect, Inadequate or Imprecise Result or Readings	23/04/2024	23/04/2024
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	23/04/2024	23/04/2024
									A0908	Incorrect, Inadequate or Imprecise Result or Readings	23/04/2024	23/04/2024
17728	2024	All other reportable incidents	17882	Physiologic-monitoring defibrillation system	CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070903	Failure to Analyze Signal		
									A070908	Failure to Sense		
							F26	No Health Consequences or Impact	A070903	Failure to Analyze Signal		
									A070908	Failure to Sense		
17729	2024	All other reportable incidents	47270	Cardiac resynchronization therapy implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E210401	Shock from Patient Lead(s)	A0709	Device Sensing Problem		
									A071302	Inappropriate/Inadequate Shock/Stimulation		
							F08	Hospitalization or Prolonged Hospitalization	A0709	Device Sensing Problem		
									A071302	Inappropriate/Inadequate Shock/Stimulation		
							F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A0709	Device Sensing Problem		
									A071302	Inappropriate/Inadequate Shock/Stimulation		
							F1203	Life Threatening Illness or Injury	A0709	Device Sensing Problem		
									A071302	Inappropriate/Inadequate Shock/Stimulation		
							F2307	Reprogramming of Device	A0709	Device Sensing Problem		
									A071302	Inappropriate/Inadequate Shock/Stimulation		
17730	2024	All other reportable incidents	35224	External pacemaker, epicardial pacing	CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A070910	Undersensing		
							F26	No Health Consequences or Impact	A070910	Undersensing		
17731	2023	All other reportable incidents	47265	Dual-chamber implantable pacemaker, rate-responsive	CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	31/08/2023	31/08/2023
							F1901	Additional Surgery	A070504	Premature Discharge of Battery	31/08/2023	31/08/2023
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	31/08/2023	31/08/2023
							F1901	Additional Surgery	A070504	Premature Discharge of Battery	31/08/2023	31/08/2023
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	31/08/2023	31/08/2023
							F1901	Additional Surgery	A070504	Premature Discharge of Battery	31/08/2023	31/08/2023
17732	2024	All other reportable incidents	47265	Dual-chamber implantable pacemaker, rate-responsive	CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A070909	Oversensing	27/10/2023	27/10/2023
									A0712	Pacing Problem	27/10/2023	27/10/2023
									A090801	Signal Artifact/Noise	27/10/2023	27/10/2023
							F1001	Absence of Treatment	A070909	Oversensing	27/10/2023	27/10/2023
									A0712	Pacing Problem	27/10/2023	27/10/2023
									A090801	Signal Artifact/Noise	27/10/2023	27/10/2023
							F1908	Prolonged Surgery	A070909	Oversensing	27/10/2023	27/10/2023
									A0712	Pacing Problem	27/10/2023	27/10/2023
									A090801	Signal Artifact/Noise	27/10/2023	27/10/2023
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A070909	Oversensing	27/10/2023	27/10/2023
									A0712	Pacing Problem	27/10/2023	27/10/2023
									A090801	Signal Artifact/Noise	27/10/2023	27/10/2023
							F1001	Absence of Treatment	A070909	Oversensing	27/10/2023	27/10/2023

									A0712	Pacing Problem	27/10/2023	27/10/2023
									A090801	Signal Artifact/Noise	27/10/2023	27/10/2023
							F1908	Prolonged Surgery	A070909	Oversensing	27/10/2023	27/10/2023
									A0712	Pacing Problem	27/10/2023	27/10/2023
									A090801	Signal Artifact/Noise	27/10/2023	27/10/2023
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A070909	Oversensing	27/10/2023	27/10/2023
									A0712	Pacing Problem	27/10/2023	27/10/2023
									A090801	Signal Artifact/Noise	27/10/2023	27/10/2023
							F1001	Absence of Treatment	A070909	Oversensing	27/10/2023	27/10/2023
									A0712	Pacing Problem	27/10/2023	27/10/2023
									A090801	Signal Artifact/Noise	27/10/2023	27/10/2023
							F1908	Prolonged Surgery	A070909	Oversensing	27/10/2023	27/10/2023
									A0712	Pacing Problem	27/10/2023	27/10/2023
									A090801	Signal Artifact/Noise	27/10/2023	27/10/2023
17733	2024	All other reportable incidents	47265	Dual-chamber implantable pacemaker, rate-responsive	CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	17/07/2024	17/07/2024
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	17/07/2024	17/07/2024
							F1903	Device Explantation	A070504	Premature Discharge of Battery	17/07/2024	17/07/2024
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	17/07/2024	17/07/2024
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	17/07/2024	17/07/2024
							F1903	Device Explantation	A070504	Premature Discharge of Battery	17/07/2024	17/07/2024
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	17/07/2024	17/07/2024
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	17/07/2024	17/07/2024
							F1903	Device Explantation	A070504	Premature Discharge of Battery	17/07/2024	17/07/2024
17734	2025		37265	Dual-chamber implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A0723	Unexpected Mode Switch		
									A0909	Unexpected Therapeutic Results		
							F01	Change in Therapeutic Response	A0723	Unexpected Mode Switch		
									A0909	Unexpected Therapeutic Results		
17735	2022	All other reportable incidents	47265	Dual-chamber implantable pacemaker, rate-responsive	CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	24/07/2023	24/07/2023
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	24/07/2023	24/07/2023
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	24/07/2023	24/07/2023
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	24/07/2023	24/07/2023
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	24/07/2023	24/07/2023
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	24/07/2023	24/07/2023
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	24/07/2023	24/07/2023
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	24/07/2023	24/07/2023
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	24/07/2023	24/07/2023
17736	2023	All other reportable incidents	47263	Cardiac resynchronization therapy implantable pacemaker	CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	13/12/2022	13/12/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	13/12/2022	13/12/2022
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	13/12/2022	13/12/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	13/12/2022	13/12/2022
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	13/12/2022	13/12/2022
							F1901	Additional Surgery	A24	No Apparent Device Problem	13/12/2022	13/12/2022
17737	2023	All other reportable incidents	47265	Dual-chamber implantable pacemaker, rate-responsive	CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
17738	2021	All other reportable incidents	35852	Single-chamber implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A070909	Oversensing	05/01/2022	05/01/2022
									A071302	Inappropriate/Inadequate Shock/Stimulation	05/01/2022	05/01/2022
							F08	Hospitalization or Prolonged Hospitalization	A070909	Oversensing	05/01/2022	05/01/2022
									A071302	Inappropriate/Inadequate Shock/Stimulation	05/01/2022	05/01/2022
							F1903	Device Explantation	A070909	Oversensing	05/01/2022	05/01/2022
									A071302	Inappropriate/Inadequate Shock/Stimulation	05/01/2022	05/01/2022
17739	2023		47265	Dual-chamber implantable pacemaker, rate-responsive	CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F1901	Additional Surgery	A070504	Premature Discharge of Battery		
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F1901	Additional Surgery	A070504	Premature Discharge of Battery		
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F1901	Additional Surgery	A070504	Premature Discharge of Battery		
17740	2020	All other reportable incidents	47265	Dual-chamber implantable pacemaker, rate-responsive	CT1685	Cardiac pacemakers	E0601	Arrhythmia	A070910	Undersensing		
									A0712	Pacing Problem		
							F23	Unexpected Medical Intervention	A070910	Undersensing		
									A0712	Pacing Problem		
					CT257	Implantable cardiac pacemakers	E0601	Arrhythmia	A070910	Undersensing		
									A0712	Pacing Problem		
							F23	Unexpected Medical Intervention	A070910	Undersensing		
									A0712	Pacing Problem		
					CT2866	Implantable cardiac electrical stimulators	E0601	Arrhythmia	A070910	Undersensing		
									A0712	Pacing Problem		
							F23	Unexpected Medical Intervention	A070910	Undersensing		
									A0712	Pacing Problem		
17741	2025	All other reportable incidents	35852	Single-chamber implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A070910	Undersensing		
									A0712	Pacing Problem		
							F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A070910	Undersensing		
									A0712	Pacing Problem		
17742	2023	All other reportable incidents	63188	Left-ventricular wireless pacing system sensor/transmitter	CT145	Cardiovascular devices			A0708	Power Problem		
					CT1577	Wireless monitoring system transmitters			A0708	Power Problem		
					CT1684	Cardiac pacemakers and associated devices			A0708	Power Problem		
					CT2727	Active implant devices/ancillaries			A0708	Power Problem		
					CT2750	Implantable cardiac electrical stimulators and associated devices			A0708	Power Problem		

					CT2885	Physiological monitoring/measuring system transmitters/receivers/transceivers			A0708	Power Problem		
17743	2023	All other reportable incidents	47205	Cardiac pulse generator programmer	CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A0203	Defective Device		
									A150105	Key or Button Unresponsive/not Working		
							F26	No Health Consequences or Impact	A0203	Defective Device		
									A150105	Key or Button Unresponsive/not Working		
17744	2021	All other reportable incidents	47267	Single-chamber implantable pacemaker, rate-responsive	CT1685	Cardiac pacemakers	E060101	Asystole	A070101	Failure to Capture	30/11/2023	30/11/2023
									A0705	Battery Problem	30/11/2023	30/11/2023
									A070504	Premature Discharge of Battery	30/11/2023	30/11/2023
									A070909	Oversensing	30/11/2023	30/11/2023
									A0908	Incorrect, Inadequate or Imprecise Result or Readings	30/11/2023	30/11/2023
							F08	Hospitalization or Prolonged Hospitalization	A070101	Failure to Capture	30/11/2023	30/11/2023
									A0705	Battery Problem	30/11/2023	30/11/2023
									A070504	Premature Discharge of Battery	30/11/2023	30/11/2023
									A070909	Oversensing	30/11/2023	30/11/2023
									A0908	Incorrect, Inadequate or Imprecise Result or Readings	30/11/2023	30/11/2023
							F1905	Device Revision or Replacement	A070101	Failure to Capture	30/11/2023	30/11/2023
									A0705	Battery Problem	30/11/2023	30/11/2023
									A070504	Premature Discharge of Battery	30/11/2023	30/11/2023
									A070909	Oversensing	30/11/2023	30/11/2023
									A0908	Incorrect, Inadequate or Imprecise Result or Readings	30/11/2023	30/11/2023
					CT257	Implantable cardiac pacemakers	E060101	Asystole	A070101	Failure to Capture	30/11/2023	30/11/2023
									A0705	Battery Problem	30/11/2023	30/11/2023
									A070504	Premature Discharge of Battery	30/11/2023	30/11/2023
									A070909	Oversensing	30/11/2023	30/11/2023
									A0908	Incorrect, Inadequate or Imprecise Result or Readings	30/11/2023	30/11/2023
							F08	Hospitalization or Prolonged Hospitalization	A070101	Failure to Capture	30/11/2023	30/11/2023
									A0705	Battery Problem	30/11/2023	30/11/2023
									A070504	Premature Discharge of Battery	30/11/2023	30/11/2023
									A070909	Oversensing	30/11/2023	30/11/2023
									A0908	Incorrect, Inadequate or Imprecise Result or Readings	30/11/2023	30/11/2023
							F1905	Device Revision or Replacement	A070101	Failure to Capture	30/11/2023	30/11/2023
									A0705	Battery Problem	30/11/2023	30/11/2023
									A070504	Premature Discharge of Battery	30/11/2023	30/11/2023
									A070909	Oversensing	30/11/2023	30/11/2023
									A0908	Incorrect, Inadequate or Imprecise Result or Readings	30/11/2023	30/11/2023
					CT2866	Implantable cardiac electrical stimulators	E060101	Asystole	A070101	Failure to Capture	30/11/2023	30/11/2023
									A0705	Battery Problem	30/11/2023	30/11/2023
									A070504	Premature Discharge of Battery	30/11/2023	30/11/2023
									A070909	Oversensing	30/11/2023	30/11/2023
									A0908	Incorrect, Inadequate or Imprecise Result or Readings	30/11/2023	30/11/2023
							F08	Hospitalization or Prolonged Hospitalization	A070101	Failure to Capture	30/11/2023	30/11/2023
									A0705	Battery Problem	30/11/2023	30/11/2023
									A070504	Premature Discharge of Battery	30/11/2023	30/11/2023
									A070909	Oversensing	30/11/2023	30/11/2023
									A0908	Incorrect, Inadequate or Imprecise Result or Readings	30/11/2023	30/11/2023
							F1905	Device Revision or Replacement	A070101	Failure to Capture	30/11/2023	30/11/2023
									A0705	Battery Problem	30/11/2023	30/11/2023
									A070504	Premature Discharge of Battery	30/11/2023	30/11/2023
									A070909	Oversensing	30/11/2023	30/11/2023
									A0908	Incorrect, Inadequate or Imprecise Result or Readings	30/11/2023	30/11/2023
17745	2022	All other reportable incidents	47267	Single-chamber implantable pacemaker, rate-responsive	CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A0711	Interrogation Problem	01/12/2025	01/12/2025
									A0908	Incorrect, Inadequate or Imprecise Result or Readings	01/12/2025	01/12/2025
							F08	Hospitalization or Prolonged Hospitalization	A0711	Interrogation Problem	01/12/2025	01/12/2025
									A0908	Incorrect, Inadequate or Imprecise Result or Readings	01/12/2025	01/12/2025
							F1905	Device Revision or Replacement	A0711	Interrogation Problem	01/12/2025	01/12/2025
									A0908	Incorrect, Inadequate or Imprecise Result or Readings	01/12/2025	01/12/2025
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A0711	Interrogation Problem	01/12/2025	01/12/2025
									A0908	Incorrect, Inadequate or Imprecise Result or Readings	01/12/2025	01/12/2025
							F08	Hospitalization or Prolonged Hospitalization	A0711	Interrogation Problem	01/12/2025	01/12/2025
									A0908	Incorrect, Inadequate or Imprecise Result or Readings	01/12/2025	01/12/2025
							F1905	Device Revision or Replacement	A0711	Interrogation Problem	01/12/2025	01/12/2025
									A0908	Incorrect, Inadequate or Imprecise Result or Readings	01/12/2025	01/12/2025
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A0711	Interrogation Problem	01/12/2025	01/12/2025
									A0908	Incorrect, Inadequate or Imprecise Result or Readings	01/12/2025	01/12/2025
							F08	Hospitalization or Prolonged Hospitalization	A0711	Interrogation Problem	01/12/2025	01/12/2025
									A0908	Incorrect, Inadequate or Imprecise Result or Readings	01/12/2025	01/12/2025
							F1905	Device Revision or Replacement	A0711	Interrogation Problem	01/12/2025	01/12/2025
									A0908	Incorrect, Inadequate or Imprecise Result or Readings	01/12/2025	01/12/2025
17746	2023	Unanticipated serious deterioration in state of health	47265	Dual-chamber implantable pacemaker, rate-responsive	CT1685	Cardiac pacemakers	E0717	Dyspnea	A0722	Impedance Problem		
									A072201	High Impedance		
									A1206	Misconnection		
							F12	Serious Injury/ Illness/ Impairment	A0722	Impedance Problem		
									A072201	High Impedance		

									A1206	Misconnection		
							F2303	Medication Required	A0722	Impedance Problem		
									A072201	High Impedance		
									A1206	Misconnection		
					CT257	Implantable cardiac pacemakers	E0717	Dyspnea	A0722	Impedance Problem		
									A072201	High Impedance		
									A1206	Misconnection		
							F12	Serious Injury/ Illness/ Impairment	A0722	Impedance Problem		
									A072201	High Impedance		
									A1206	Misconnection		
							F2303	Medication Required	A0722	Impedance Problem		
									A072201	High Impedance		
									A1206	Misconnection		
					CT2866	Implantable cardiac electrical stimulators	E0717	Dyspnea	A0722	Impedance Problem		
									A072201	High Impedance		
									A1206	Misconnection		
							F12	Serious Injury/ Illness/ Impairment	A0722	Impedance Problem		
									A072201	High Impedance		
									A1206	Misconnection		
							F2303	Medication Required	A0722	Impedance Problem		
									A072201	High Impedance		
									A1206	Misconnection		
17747	2023		47265	Dual-chamber implantable pacemaker, rate-responsive	CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem		
							F1901	Additional Surgery	A24	No Apparent Device Problem		
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem		
							F1901	Additional Surgery	A24	No Apparent Device Problem		
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem		
							F1901	Additional Surgery	A24	No Apparent Device Problem		
17748	2023		47265	Dual-chamber implantable pacemaker, rate-responsive	CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	28/03/2023	28/03/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	28/03/2023	28/03/2023
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	28/03/2023	28/03/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	28/03/2023	28/03/2023
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	28/03/2023	28/03/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	28/03/2023	28/03/2023
17749	2020	All other reportable incidents	60789	Intracardiac pacemaker	CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F22	Unexpected Diagnostic Intervention	A070504	Premature Discharge of Battery		
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F22	Unexpected Diagnostic Intervention	A070504	Premature Discharge of Battery		
					CT2818	Catheterization/Cannulation/Fluid path devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F22	Unexpected Diagnostic Intervention	A070504	Premature Discharge of Battery		
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F22	Unexpected Diagnostic Intervention	A070504	Premature Discharge of Battery		
17750	2020	All other reportable incidents	36241	Pacemaker/defibrillator lead	CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A05	Mechanical Problem	11/03/2020	11/03/2020
							F26	No Health Consequences or Impact	A05	Mechanical Problem	11/03/2020	11/03/2020
17751	2023	All other reportable incidents	35852	Single-chamber implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E2104	Electric Shock	A0709	Device Sensing Problem	24/04/2023	24/04/2023
									A070909	Oversensing	24/04/2023	24/04/2023
									A071302	Inappropriate/Inadequate Shock/Stimulation	24/04/2023	24/04/2023
									A090801	Signal Artifact/Noise	24/04/2023	24/04/2023
									A1301	Failure to Read Input Signal	24/04/2023	24/04/2023
							F08	Hospitalization or Prolonged Hospitalization	A0709	Device Sensing Problem	24/04/2023	24/04/2023
									A070909	Oversensing	24/04/2023	24/04/2023
									A071302	Inappropriate/Inadequate Shock/Stimulation	24/04/2023	24/04/2023
									A090801	Signal Artifact/Noise	24/04/2023	24/04/2023
									A1301	Failure to Read Input Signal	24/04/2023	24/04/2023
							F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A0709	Device Sensing Problem	24/04/2023	24/04/2023
									A070909	Oversensing	24/04/2023	24/04/2023
									A071302	Inappropriate/Inadequate Shock/Stimulation	24/04/2023	24/04/2023
									A090801	Signal Artifact/Noise	24/04/2023	24/04/2023
									A1301	Failure to Read Input Signal	24/04/2023	24/04/2023
							F1905	Device Revision or Replacement	A0709	Device Sensing Problem	24/04/2023	24/04/2023
									A070909	Oversensing	24/04/2023	24/04/2023
									A071302	Inappropriate/Inadequate Shock/Stimulation	24/04/2023	24/04/2023
									A090801	Signal Artifact/Noise	24/04/2023	24/04/2023
									A1301	Failure to Read Input Signal	24/04/2023	24/04/2023
							F2203	Imaging Required	A0709	Device Sensing Problem	24/04/2023	24/04/2023
									A070909	Oversensing	24/04/2023	24/04/2023
									A071302	Inappropriate/Inadequate Shock/Stimulation	24/04/2023	24/04/2023
									A090801	Signal Artifact/Noise	24/04/2023	24/04/2023
									A1301	Failure to Read Input Signal	24/04/2023	24/04/2023
17752	2022	All other reportable incidents	47263	Cardiac resynchronization therapy implantable pacemaker	CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A0908	Incorrect, Inadequate or Imprecise Result or Readings		
							F08	Hospitalization or Prolonged Hospitalization	A0908	Incorrect, Inadequate or Imprecise Result or Readings		
							F1905	Device Revision or Replacement	A0908	Incorrect, Inadequate or Imprecise Result or Readings		
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A0908	Incorrect, Inadequate or Imprecise Result or Readings		
							F08	Hospitalization or Prolonged Hospitalization	A0908	Incorrect, Inadequate or Imprecise Result or Readings		
							F1905	Device Revision or Replacement	A0908	Incorrect, Inadequate or Imprecise Result or Readings		
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A0908	Incorrect, Inadequate or Imprecise Result or Readings		
							F08	Hospitalization or Prolonged Hospitalization	A0908	Incorrect, Inadequate or Imprecise Result or Readings		
							F1905	Device Revision or Replacement	A0908	Incorrect, Inadequate or Imprecise Result or Readings		
17753	2024	All other reportable incidents	35852	Single-chamber implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E060102	Atrial Fibrillation	A27	Appropriate Device Problem Term/Code not Available		

							F12	Serious Injury/ Illness/ Impairment	A27	Appropriate Device Problem Term/Code not Available		
							F23	Unexpected Medical Intervention	A27	Appropriate Device Problem Term/Code not Available		
17754	2024	All other reportable incidents	47265	Dual-chamber implantable pacemaker, rate-responsive	CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A071102	Failure to Interrogate	27/06/2024	27/06/2024
									A071204	No Pacing	27/06/2024	27/06/2024
							F1901	Additional Surgery	A071102	Failure to Interrogate	27/06/2024	27/06/2024
									A071204	No Pacing	27/06/2024	27/06/2024
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A071102	Failure to Interrogate	27/06/2024	27/06/2024
									A071204	No Pacing	27/06/2024	27/06/2024
							F1901	Additional Surgery	A071102	Failure to Interrogate	27/06/2024	27/06/2024
									A071204	No Pacing	27/06/2024	27/06/2024
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A071102	Failure to Interrogate	27/06/2024	27/06/2024
									A071204	No Pacing	27/06/2024	27/06/2024
							F1901	Additional Surgery	A071102	Failure to Interrogate	27/06/2024	27/06/2024
									A071204	No Pacing	27/06/2024	27/06/2024
17755	2024	All other reportable incidents	47265	Dual-chamber implantable pacemaker, rate-responsive	CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A0908	Incorrect, Inadequate or Imprecise Result or Readings		
							F26	No Health Consequences or Impact	A0908	Incorrect, Inadequate or Imprecise Result or Readings		
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A0908	Incorrect, Inadequate or Imprecise Result or Readings		
							F26	No Health Consequences or Impact	A0908	Incorrect, Inadequate or Imprecise Result or Readings		
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A0908	Incorrect, Inadequate or Imprecise Result or Readings		
							F26	No Health Consequences or Impact	A0908	Incorrect, Inadequate or Imprecise Result or Readings		
17756	2023	All other reportable incidents	35852	Single-chamber implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	11/02/2024	11/02/2024
									A0908	Incorrect, Inadequate or Imprecise Result or Readings	11/02/2024	11/02/2024
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	11/02/2024	11/02/2024
									A0908	Incorrect, Inadequate or Imprecise Result or Readings	11/02/2024	11/02/2024
							F1903	Device Explantation	A070504	Premature Discharge of Battery	11/02/2024	11/02/2024
									A0908	Incorrect, Inadequate or Imprecise Result or Readings	11/02/2024	11/02/2024
17757	2024	Death	47263	Cardiac resynchronization therapy implantable pacemaker	CT1685	Cardiac pacemakers	E060110	Ventricular Fibrillation	A27	Appropriate Device Problem Term/Code not Available		
							E2401	Insufficient Information	A27	Appropriate Device Problem Term/Code not Available		
							F02	Death	A27	Appropriate Device Problem Term/Code not Available		
							F2306	Resuscitation	A27	Appropriate Device Problem Term/Code not Available		
					CT257	Implantable cardiac pacemakers	E060110	Ventricular Fibrillation	A27	Appropriate Device Problem Term/Code not Available		
							E2401	Insufficient Information	A27	Appropriate Device Problem Term/Code not Available		
							F02	Death	A27	Appropriate Device Problem Term/Code not Available		
							F2306	Resuscitation	A27	Appropriate Device Problem Term/Code not Available		
					CT2866	Implantable cardiac electrical stimulators	E060110	Ventricular Fibrillation	A27	Appropriate Device Problem Term/Code not Available		
							E2401	Insufficient Information	A27	Appropriate Device Problem Term/Code not Available		
							F02	Death	A27	Appropriate Device Problem Term/Code not Available		
							F2306	Resuscitation	A27	Appropriate Device Problem Term/Code not Available		
17758	2024	All other reportable incidents	47265	Dual-chamber implantable pacemaker, rate-responsive	CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A071102	Failure to Interrogate	06/06/2024	06/06/2024
									A071204	No Pacing	06/06/2024	06/06/2024
							F1901	Additional Surgery	A071102	Failure to Interrogate	06/06/2024	06/06/2024
									A071204	No Pacing	06/06/2024	06/06/2024
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A071102	Failure to Interrogate	06/06/2024	06/06/2024
									A071204	No Pacing	06/06/2024	06/06/2024
							F1901	Additional Surgery	A071102	Failure to Interrogate	06/06/2024	06/06/2024
									A071204	No Pacing	06/06/2024	06/06/2024
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A071102	Failure to Interrogate	06/06/2024	06/06/2024
									A071204	No Pacing	06/06/2024	06/06/2024
							F1901	Additional Surgery	A071102	Failure to Interrogate	06/06/2024	06/06/2024
									A071204	No Pacing	06/06/2024	06/06/2024
17759	2023	All other reportable incidents	47265	Dual-chamber implantable pacemaker, rate-responsive	CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
17760	2023	All other reportable incidents	47265	Dual-chamber implantable pacemaker, rate-responsive	CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		

					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
17761	2024	All other reportable incidents	47265	Dual-chamber implantable pacemaker, rate-responsive	CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem	29/04/2024	29/04/2024
									A0908	Incorrect, Inadequate or Imprecise Result or Readings	29/04/2024	29/04/2024
							F08	Hospitalization or Prolonged Hospitalization	A0705	Battery Problem	29/04/2024	29/04/2024
									A0908	Incorrect, Inadequate or Imprecise Result or Readings	29/04/2024	29/04/2024
							F1905	Device Revision or Replacement	A0705	Battery Problem	29/04/2024	29/04/2024
									A0908	Incorrect, Inadequate or Imprecise Result or Readings	29/04/2024	29/04/2024
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem	29/04/2024	29/04/2024
									A0908	Incorrect, Inadequate or Imprecise Result or Readings	29/04/2024	29/04/2024
							F08	Hospitalization or Prolonged Hospitalization	A0705	Battery Problem	29/04/2024	29/04/2024
									A0908	Incorrect, Inadequate or Imprecise Result or Readings	29/04/2024	29/04/2024
							F1905	Device Revision or Replacement	A0705	Battery Problem	29/04/2024	29/04/2024
									A0908	Incorrect, Inadequate or Imprecise Result or Readings	29/04/2024	29/04/2024
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem	29/04/2024	29/04/2024
									A0908	Incorrect, Inadequate or Imprecise Result or Readings	29/04/2024	29/04/2024
							F08	Hospitalization or Prolonged Hospitalization	A0705	Battery Problem	29/04/2024	29/04/2024
									A0908	Incorrect, Inadequate or Imprecise Result or Readings	29/04/2024	29/04/2024
							F1905	Device Revision or Replacement	A0705	Battery Problem	29/04/2024	29/04/2024
									A0908	Incorrect, Inadequate or Imprecise Result or Readings	29/04/2024	29/04/2024
17762	2024	All other reportable incidents	47265	Dual-chamber implantable pacemaker, rate-responsive	CT1685	Cardiac pacemakers	E2402	Appropriate Clinical Signs, Symptoms and Conditions Term/Code not Available	A1605	Reset Problem		
							F28	Appropriate Health Impact Term/Code not Available	A1605	Reset Problem		
					CT257	Implantable cardiac pacemakers	E2402	Appropriate Clinical Signs, Symptoms and Conditions Term/Code not Available	A1605	Reset Problem		
							F28	Appropriate Health Impact Term/Code not Available	A1605	Reset Problem		
					CT2866	Implantable cardiac electrical stimulators	E2402	Appropriate Clinical Signs, Symptoms and Conditions Term/Code not Available	A1605	Reset Problem		
							F28	Appropriate Health Impact Term/Code not Available	A1605	Reset Problem		
17763	2024	All other reportable incidents	17882	Physiologic-monitoring defibrillation system	CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A090101	Inappropriate Audible Prompt/Feedback		
							F27	Problem Identified During Non-Clinical Procedure	A090101	Inappropriate Audible Prompt/Feedback		
17764	2024	All other reportable incidents	35224	External pacemaker, epicardial pacing	CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A09	Output Problem		
							F27	Problem Identified During Non-Clinical Procedure	A09	Output Problem		
17765	2024		47265	Dual-chamber implantable pacemaker, rate-responsive	CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	12/12/2023	12/12/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	12/12/2023	12/12/2023
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	12/12/2023	12/12/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	12/12/2023	12/12/2023
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	12/12/2023	12/12/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	12/12/2023	12/12/2023
17766	2024	All other reportable incidents	35852	Single-chamber implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A072201	High Impedance		
									A1107	Data Problem		
							F26	No Health Consequences or Impact	A072201	High Impedance		
									A1107	Data Problem		
17767	2024	All other reportable incidents	35852	Single-chamber implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E210401	Shock from Patient Lead(s)	A070910	Undersensing		
									A0712	Pacing Problem		
									A071302	Inappropriate/Inadequate Shock/Stimulation		
							F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A070910	Undersensing		
									A0712	Pacing Problem		
									A071302	Inappropriate/Inadequate Shock/Stimulation		
17768	2023	All other reportable incidents	35852	Single-chamber implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	05/02/2024	05/02/2024
									A0908	Incorrect, Inadequate or Imprecise Result or Readings	05/02/2024	05/02/2024
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	05/02/2024	05/02/2024
									A0908	Incorrect, Inadequate or Imprecise Result or Readings	05/02/2024	05/02/2024
							F1903	Device Explantation	A070504	Premature Discharge of Battery	05/02/2024	05/02/2024
									A0908	Incorrect, Inadequate or Imprecise Result or Readings	05/02/2024	05/02/2024
17769	2024	All other reportable incidents	47265	Dual-chamber implantable pacemaker, rate-responsive	CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	16/11/2023	16/11/2023
									A071102	Failure to Interrogate	16/11/2023	16/11/2023
							F1901	Additional Surgery	A070504	Premature Discharge of Battery	16/11/2023	16/11/2023
									A071102	Failure to Interrogate	16/11/2023	16/11/2023
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	16/11/2023	16/11/2023
									A071102	Failure to Interrogate	16/11/2023	16/11/2023
							F1901	Additional Surgery	A070504	Premature Discharge of Battery	16/11/2023	16/11/2023
									A071102	Failure to Interrogate	16/11/2023	16/11/2023
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	16/11/2023	16/11/2023
									A071102	Failure to Interrogate	16/11/2023	16/11/2023
							F1901	Additional Surgery	A070504	Premature Discharge of Battery	16/11/2023	16/11/2023
									A071102	Failure to Interrogate	16/11/2023	16/11/2023
17770	2024		47270	Cardiac resynchronization therapy implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A0202	Defective Component		
									A0203	Defective Device		
									A0705	Battery Problem		
							F1903	Device Explantation	A0202	Defective Component		
									A0203	Defective Device		
									A0705	Battery Problem		
							F26	No Health Consequences or Impact	A0202	Defective Component		
									A0203	Defective Device		
									A0705	Battery Problem		
17771	2024	All other reportable incidents	47263	Cardiac resynchronization therapy implantable pacemaker	CT1685	Cardiac pacemakers	E2103	Device Overstimulation of Tissue	A0908	Incorrect, Inadequate or Imprecise Result or Readings	24/03/2024	24/03/2024

							F08	Hospitalization or Prolonged Hospitalization	A0908	Incorrect, Inadequate or Imprecise Result or Readings	24/03/2024	24/03/2024
							F1903	Device Explantation	A0908	Incorrect, Inadequate or Imprecise Result or Readings	24/03/2024	24/03/2024
					CT257	Implantable cardiac pacemakers	E2103	Device Overstimulation of Tissue	A0908	Incorrect, Inadequate or Imprecise Result or Readings	24/03/2024	24/03/2024
							F08	Hospitalization or Prolonged Hospitalization	A0908	Incorrect, Inadequate or Imprecise Result or Readings	24/03/2024	24/03/2024
							F1903	Device Explantation	A0908	Incorrect, Inadequate or Imprecise Result or Readings	24/03/2024	24/03/2024
					CT2866	Implantable cardiac electrical stimulators	E2103	Device Overstimulation of Tissue	A0908	Incorrect, Inadequate or Imprecise Result or Readings	24/03/2024	24/03/2024
							F08	Hospitalization or Prolonged Hospitalization	A0908	Incorrect, Inadequate or Imprecise Result or Readings	24/03/2024	24/03/2024
							F1903	Device Explantation	A0908	Incorrect, Inadequate or Imprecise Result or Readings	24/03/2024	24/03/2024
17772	2024	All other reportable incidents	47270	Cardiac resynchronization therapy implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E060109	Tachycardia	A0701	Capturing Problem		
							E060110	Ventricular Fibrillation	A0701	Capturing Problem		
							F1203	Life Threatening Illness or Injury	A0701	Capturing Problem		
17773	2024	All other reportable incidents	47265	Dual-chamber implantable pacemaker, rate-responsive	CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
17774	2024		47265	Dual-chamber implantable pacemaker, rate-responsive	CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	11/03/2024	11/03/2024
							F1901	Additional Surgery	A24	No Apparent Device Problem	11/03/2024	11/03/2024
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	11/03/2024	11/03/2024
							F1901	Additional Surgery	A24	No Apparent Device Problem	11/03/2024	11/03/2024
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	11/03/2024	11/03/2024
							F1901	Additional Surgery	A24	No Apparent Device Problem	11/03/2024	11/03/2024
17775	2024	All other reportable incidents	47270	Cardiac resynchronization therapy implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	04/02/2025	04/02/2025
									A0908	Incorrect, Inadequate or Imprecise Result or Readings	04/02/2025	04/02/2025
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	04/02/2025	04/02/2025
									A0908	Incorrect, Inadequate or Imprecise Result or Readings	04/02/2025	04/02/2025
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	04/02/2025	04/02/2025
									A0908	Incorrect, Inadequate or Imprecise Result or Readings	04/02/2025	04/02/2025
17776	2024	All other reportable incidents	47265	Dual-chamber implantable pacemaker, rate-responsive	CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem	22/08/2024	22/08/2024
									A070504	Premature Discharge of Battery	22/08/2024	22/08/2024
							F08	Hospitalization or Prolonged Hospitalization	A0705	Battery Problem	22/08/2024	22/08/2024
									A070504	Premature Discharge of Battery	22/08/2024	22/08/2024
							F1905	Device Revision or Replacement	A0705	Battery Problem	22/08/2024	22/08/2024
									A070504	Premature Discharge of Battery	22/08/2024	22/08/2024
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem	22/08/2024	22/08/2024
									A070504	Premature Discharge of Battery	22/08/2024	22/08/2024
							F08	Hospitalization or Prolonged Hospitalization	A0705	Battery Problem	22/08/2024	22/08/2024
									A070504	Premature Discharge of Battery	22/08/2024	22/08/2024
							F1905	Device Revision or Replacement	A0705	Battery Problem	22/08/2024	22/08/2024
									A070504	Premature Discharge of Battery	22/08/2024	22/08/2024
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem	22/08/2024	22/08/2024
									A070504	Premature Discharge of Battery	22/08/2024	22/08/2024
							F08	Hospitalization or Prolonged Hospitalization	A0705	Battery Problem	22/08/2024	22/08/2024
									A070504	Premature Discharge of Battery	22/08/2024	22/08/2024
							F1905	Device Revision or Replacement	A0705	Battery Problem	22/08/2024	22/08/2024
									A070504	Premature Discharge of Battery	22/08/2024	22/08/2024
17777	2023	All other reportable incidents	17882	Physiologic-monitoring defibrillation system	CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A07	Electrical /Electronic Property Problem		
							F27	Problem Identified During Non-Clinical Procedure	A07	Electrical /Electronic Property Problem		
17778	2024	All other reportable incidents	35224	External pacemaker, epicardial pacing	CT1685	Cardiac pacemakers	E060110	Ventricular Fibrillation	A0709	Device Sensing Problem		
									A09	Output Problem		
							F1203	Life Threatening Illness or Injury	A0709	Device Sensing Problem		
									A09	Output Problem		
							F23	Unexpected Medical Intervention	A0709	Device Sensing Problem		
									A09	Output Problem		
17779	2024	All other reportable incidents	47265	Dual-chamber implantable pacemaker, rate-responsive	CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
17780	2024	All other reportable incidents	17882	Physiologic-monitoring defibrillation system	CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A071301	Failure to Deliver Shock/Stimulation		
							F26	No Health Consequences or Impact	A071301	Failure to Deliver Shock/Stimulation		
17781	2023		47265	Dual-chamber implantable pacemaker, rate-responsive	CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem		
							F1901	Additional Surgery	A24	No Apparent Device Problem		
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem		
							F1901	Additional Surgery	A24	No Apparent Device Problem		
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem		
							F1901	Additional Surgery	A24	No Apparent Device Problem		
17782	2023		47265	Dual-chamber implantable pacemaker, rate-responsive	CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem		
							F1901	Additional Surgery	A24	No Apparent Device Problem		
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem		
							F1901	Additional Surgery	A24	No Apparent Device Problem		
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem		
							F1901	Additional Surgery	A24	No Apparent Device Problem		
17783	2023	All other reportable incidents	35852	Single-chamber implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E2104	Electric Shock	A0709	Device Sensing Problem		
									A070909	Oversensing		
									A071302	Inappropriate/Inadequate Shock/Stimulation		
									A1301	Failure to Read Input Signal		

							F08	Hospitalization or Prolonged Hospitalization	A0709	Device Sensing Problem		
									A070909	Oversensing		
									A071302	Inappropriate/Inadequate Shock/Stimulation		
									A1301	Failure to Read Input Signal		
							F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A0709	Device Sensing Problem		
									A070909	Oversensing		
									A071302	Inappropriate/Inadequate Shock/Stimulation		
									A1301	Failure to Read Input Signal		
17784	2022	All other reportable incidents	47263	Cardiac resynchronization therapy implantable pacemaker	CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A0908	Incorrect, Inadequate or Imprecise Result or Readings	25/08/2022	25/08/2022
									A160503	Inappropriate or Unexpected Reset	25/08/2022	25/08/2022
							F08	Hospitalization or Prolonged Hospitalization	A0908	Incorrect, Inadequate or Imprecise Result or Readings	25/08/2022	25/08/2022
									A160503	Inappropriate or Unexpected Reset	25/08/2022	25/08/2022
							F1905	Device Revision or Replacement	A0908	Incorrect, Inadequate or Imprecise Result or Readings	25/08/2022	25/08/2022
									A160503	Inappropriate or Unexpected Reset	25/08/2022	25/08/2022
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A0908	Incorrect, Inadequate or Imprecise Result or Readings	25/08/2022	25/08/2022
									A160503	Inappropriate or Unexpected Reset	25/08/2022	25/08/2022
							F08	Hospitalization or Prolonged Hospitalization	A0908	Incorrect, Inadequate or Imprecise Result or Readings	25/08/2022	25/08/2022
									A160503	Inappropriate or Unexpected Reset	25/08/2022	25/08/2022
							F1905	Device Revision or Replacement	A0908	Incorrect, Inadequate or Imprecise Result or Readings	25/08/2022	25/08/2022
									A160503	Inappropriate or Unexpected Reset	25/08/2022	25/08/2022
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A0908	Incorrect, Inadequate or Imprecise Result or Readings	25/08/2022	25/08/2022
									A160503	Inappropriate or Unexpected Reset	25/08/2022	25/08/2022
							F08	Hospitalization or Prolonged Hospitalization	A0908	Incorrect, Inadequate or Imprecise Result or Readings	25/08/2022	25/08/2022
									A160503	Inappropriate or Unexpected Reset	25/08/2022	25/08/2022
							F1905	Device Revision or Replacement	A0908	Incorrect, Inadequate or Imprecise Result or Readings	25/08/2022	25/08/2022
									A160503	Inappropriate or Unexpected Reset	25/08/2022	25/08/2022
17785	2023		47265	Dual-chamber implantable pacemaker, rate-responsive	CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	02/10/2023	02/10/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	02/10/2023	02/10/2023
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	02/10/2023	02/10/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	02/10/2023	02/10/2023
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	02/10/2023	02/10/2023
							F1901	Additional Surgery	A24	No Apparent Device Problem	02/10/2023	02/10/2023
17786	2022	All other reportable incidents	47263	Cardiac resynchronization therapy implantable pacemaker	CT1685	Cardiac pacemakers	E0112	Dizziness	A070909	Oversensing		
							F23	Unexpected Medical Intervention	A070909	Oversensing		
					CT257	Implantable cardiac pacemakers	E0112	Dizziness	A070909	Oversensing		
							F23	Unexpected Medical Intervention	A070909	Oversensing		
					CT2866	Implantable cardiac electrical stimulators	E0112	Dizziness	A070909	Oversensing		
							F23	Unexpected Medical Intervention	A070909	Oversensing		
17787	2024	All other reportable incidents	47265	Dual-chamber implantable pacemaker, rate-responsive	CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A072001	Electromagnetic Interference		
							F26	No Health Consequences or Impact	A072001	Electromagnetic Interference		
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A072001	Electromagnetic Interference		
							F26	No Health Consequences or Impact	A072001	Electromagnetic Interference		
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A072001	Electromagnetic Interference		
							F26	No Health Consequences or Impact	A072001	Electromagnetic Interference		
17788	2023	All other reportable incidents	35852	Single-chamber implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A070909	Oversensing		
									A090801	Signal Artifact/Noise		
							F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A070909	Oversensing		
									A090801	Signal Artifact/Noise		
							F2203	Imaging Required	A070909	Oversensing		
									A090801	Signal Artifact/Noise		
							F23	Unexpected Medical Intervention	A070909	Oversensing		
									A090801	Signal Artifact/Noise		
17789	2023		47265	Dual-chamber implantable pacemaker, rate-responsive	CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem		
							F1901	Additional Surgery	A24	No Apparent Device Problem		
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem		
							F1901	Additional Surgery	A24	No Apparent Device Problem		
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem		
							F1901	Additional Surgery	A24	No Apparent Device Problem		
17790	2023	All other reportable incidents	35852	Single-chamber implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A0711	Interrogation Problem		
							F26	No Health Consequences or Impact	A0711	Interrogation Problem		
17791	2023		47265	Dual-chamber implantable pacemaker, rate-responsive	CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A02	Manufacturing, Packaging or Shipping Problem		
							F1905	Device Revision or Replacement	A02	Manufacturing, Packaging or Shipping Problem		
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A02	Manufacturing, Packaging or Shipping Problem		
							F1905	Device Revision or Replacement	A02	Manufacturing, Packaging or Shipping Problem		
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A02	Manufacturing, Packaging or Shipping Problem		
							F1905	Device Revision or Replacement	A02	Manufacturing, Packaging or Shipping Problem		
17792	2024		47270	Cardiac resynchronization therapy implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F05	Delay to Treatment/ Therapy	A070504	Premature Discharge of Battery		
17793	2023	All other reportable incidents	35852	Single-chamber implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E2104	Electric Shock	A0709	Device Sensing Problem	24/04/2023	24/04/2023
									A070909	Oversensing	24/04/2023	24/04/2023
									A071302	Inappropriate/Inadequate Shock/Stimulation	24/04/2023	24/04/2023
									A090801	Signal Artifact/Noise	24/04/2023	24/04/2023
									A1301	Failure to Read Input Signal	24/04/2023	24/04/2023
							F08	Hospitalization or Prolonged Hospitalization	A0709	Device Sensing Problem	24/04/2023	24/04/2023
									A070909	Oversensing	24/04/2023	24/04/2023
									A071302	Inappropriate/Inadequate Shock/Stimulation	24/04/2023	24/04/2023
									A090801	Signal Artifact/Noise	24/04/2023	24/04/2023
									A1301	Failure to Read Input Signal	24/04/2023	24/04/2023
							F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A0709	Device Sensing Problem	24/04/2023	24/04/2023
									A070909	Oversensing	24/04/2023	24/04/2023
									A071302	Inappropriate/Inadequate Shock/Stimulation	24/04/2023	24/04/2023

									A090801	Signal Artifact/Noise	24/04/2023	24/04/2023
									A1301	Failure to Read Input Signal	24/04/2023	24/04/2023
							F1905	Device Revision or Replacement	A0709	Device Sensing Problem	24/04/2023	24/04/2023
									A070909	Oversensing	24/04/2023	24/04/2023
									A071302	Inappropriate/Inadequate Shock/Stimulation	24/04/2023	24/04/2023
									A090801	Signal Artifact/Noise	24/04/2023	24/04/2023
									A1301	Failure to Read Input Signal	24/04/2023	24/04/2023
							F2203	Imaging Required	A0709	Device Sensing Problem	24/04/2023	24/04/2023
									A070909	Oversensing	24/04/2023	24/04/2023
									A071302	Inappropriate/Inadequate Shock/Stimulation	24/04/2023	24/04/2023
									A090801	Signal Artifact/Noise	24/04/2023	24/04/2023
									A1301	Failure to Read Input Signal	24/04/2023	24/04/2023
17794	2023	All other reportable incidents	35852	Single-chamber implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E2104	Electric Shock	A0709	Device Sensing Problem		
									A070909	Oversensing		
									A071302	Inappropriate/Inadequate Shock/Stimulation		
							F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A0709	Device Sensing Problem		
									A070909	Oversensing		
									A071302	Inappropriate/Inadequate Shock/Stimulation		
17795	2024	All other reportable incidents	47270	Cardiac resynchronization therapy implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A05	Mechanical Problem	25/04/2024	25/04/2024
							F08	Hospitalization or Prolonged Hospitalization	A05	Mechanical Problem	25/04/2024	25/04/2024
							F12	Serious Injury/ Illness/ Impairment	A05	Mechanical Problem	25/04/2024	25/04/2024
							F1905	Device Revision or Replacement	A05	Mechanical Problem	25/04/2024	25/04/2024
17796	2024	All other reportable incidents	47265	Dual-chamber implantable pacemaker, rate-responsive	CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A070909	Oversensing		
									A090801	Signal Artifact/Noise		
							F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A070909	Oversensing		
									A090801	Signal Artifact/Noise		
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A070909	Oversensing		
									A090801	Signal Artifact/Noise		
							F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A070909	Oversensing		
									A090801	Signal Artifact/Noise		
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A070909	Oversensing		
									A090801	Signal Artifact/Noise		
							F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A070909	Oversensing		
									A090801	Signal Artifact/Noise		
17797	2024		47265	Dual-chamber implantable pacemaker, rate-responsive	CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem		
							F1901	Additional Surgery	A24	No Apparent Device Problem		
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem		
							F1901	Additional Surgery	A24	No Apparent Device Problem		
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem		
							F1901	Additional Surgery	A24	No Apparent Device Problem		
17798	2025		47265	Dual-chamber implantable pacemaker, rate-responsive	CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A071205	Pacemaker Found in Back-Up Mode		
									A2205	Device Provides Uninformative Message		
							F26	No Health Consequences or Impact	A071205	Pacemaker Found in Back-Up Mode		
									A2205	Device Provides Uninformative Message		
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A071205	Pacemaker Found in Back-Up Mode		
									A2205	Device Provides Uninformative Message		
							F26	No Health Consequences or Impact	A071205	Pacemaker Found in Back-Up Mode		
									A2205	Device Provides Uninformative Message		
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A071205	Pacemaker Found in Back-Up Mode		
									A2205	Device Provides Uninformative Message		
							F26	No Health Consequences or Impact	A071205	Pacemaker Found in Back-Up Mode		
									A2205	Device Provides Uninformative Message		
17799	2025	All other reportable incidents	47263	Cardiac resynchronization therapy implantable pacemaker	CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
17800	2025	All other reportable incidents	47265	Dual-chamber implantable pacemaker, rate-responsive			E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
17801	2025	All other reportable incidents	60789	Intracardiac pacemaker			E2403	No Clinical Signs, Symptoms or Conditions	A0712	Pacing Problem		
							F2307	Reprogramming of Device	A0712	Pacing Problem		
17802	2025	All other reportable incidents	60789	Intracardiac pacemaker			E2403	No Clinical Signs, Symptoms or Conditions	A070102	High Capture Threshold	21/11/2025	21/11/2025
									A070901	Decreased Sensitivity	21/11/2025	21/11/2025
									A1502	Positioning Problem	21/11/2025	21/11/2025
							F26	No Health Consequences or Impact	A070102	High Capture Threshold	21/11/2025	21/11/2025
									A070901	Decreased Sensitivity	21/11/2025	21/11/2025
									A1502	Positioning Problem	21/11/2025	21/11/2025
17803	2025	All other reportable incidents	47267	Single-chamber implantable pacemaker, rate-responsive			E1906	Unspecified Infection	A24	No Apparent Device Problem	18/11/2025	18/11/2025
							F08	Hospitalization or Prolonged Hospitalization	A24	No Apparent Device Problem	18/11/2025	18/11/2025
							F1203	Life Threatening Illness or Injury	A24	No Apparent Device Problem	18/11/2025	18/11/2025
							F1905	Device Revision or Replacement	A24	No Apparent Device Problem	18/11/2025	18/11/2025
17804	2025		47265	Dual-chamber implantable pacemaker, rate-responsive	CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A0709	Device Sensing Problem		
							F1908	Prolonged Surgery	A0709	Device Sensing Problem		
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A0709	Device Sensing Problem		
							F1908	Prolonged Surgery	A0709	Device Sensing Problem		
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A0709	Device Sensing Problem		
							F1908	Prolonged Surgery	A0709	Device Sensing Problem		
17805	2025	All other reportable incidents	47265	Dual-chamber implantable pacemaker, rate-responsive			E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
17806	2025	All other reportable incidents	47265	Dual-chamber implantable pacemaker, rate-responsive	CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A0908	Incorrect, Inadequate or Imprecise Result or Readings	19/01/2026	19/01/2026
							F08	Hospitalization or Prolonged Hospitalization	A0908	Incorrect, Inadequate or Imprecise Result or Readings	19/01/2026	19/01/2026
							F1905	Device Revision or Replacement	A0908	Incorrect, Inadequate or Imprecise Result or Readings	19/01/2026	19/01/2026
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A0908	Incorrect, Inadequate or Imprecise Result or Readings	19/01/2026	19/01/2026

							F08	Hospitalization or Prolonged Hospitalization	A0908	Incorrect, Inadequate or Imprecise Result or Readings	19/01/2026	19/01/2026
							F1905	Device Revision or Replacement	A0908	Incorrect, Inadequate or Imprecise Result or Readings	19/01/2026	19/01/2026
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A0908	Incorrect, Inadequate or Imprecise Result or Readings	19/01/2026	19/01/2026
							F08	Hospitalization or Prolonged Hospitalization	A0908	Incorrect, Inadequate or Imprecise Result or Readings	19/01/2026	19/01/2026
							F1905	Device Revision or Replacement	A0908	Incorrect, Inadequate or Imprecise Result or Readings	19/01/2026	19/01/2026
17807	2025	All other reportable incidents	47263	Cardiac resynchronization therapy implantable pacemaker	CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A070909	Oversensing		
									A0711	Interrogation Problem		
									A0908	Incorrect, Inadequate or Imprecise Result or Readings		
							F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A070909	Oversensing		
									A0711	Interrogation Problem		
									A0908	Incorrect, Inadequate or Imprecise Result or Readings		
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A070909	Oversensing		
									A0711	Interrogation Problem		
									A0908	Incorrect, Inadequate or Imprecise Result or Readings		
							F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A070909	Oversensing		
									A0711	Interrogation Problem		
									A0908	Incorrect, Inadequate or Imprecise Result or Readings		
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A070909	Oversensing		
									A0711	Interrogation Problem		
									A0908	Incorrect, Inadequate or Imprecise Result or Readings		
							F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A070909	Oversensing		
									A0711	Interrogation Problem		
									A0908	Incorrect, Inadequate or Imprecise Result or Readings		
17808	2025		47265	Dual-chamber implantable pacemaker, rate-responsive			E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	15/11/2025	15/11/2025
							F1901	Additional Surgery	A24	No Apparent Device Problem	15/11/2025	15/11/2025
17809	2025		47265	Dual-chamber implantable pacemaker, rate-responsive			E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	02/08/2025	02/08/2025
							F1901	Additional Surgery	A24	No Apparent Device Problem	02/08/2025	02/08/2025
17810	2025	All other reportable incidents	47265	Dual-chamber implantable pacemaker, rate-responsive			E2403	No Clinical Signs, Symptoms or Conditions	A071102	Failure to Interrogate	20/11/2025	20/11/2025
							F1901	Additional Surgery	A071102	Failure to Interrogate	20/11/2025	20/11/2025
17811	2025	All other reportable incidents	47265	Dual-chamber implantable pacemaker, rate-responsive			E012207	Twitching	A0908	Incorrect, Inadequate or Imprecise Result or Readings	14/11/2025	14/11/2025
							E2103	Device Overstimulation of Tissue	A0908	Incorrect, Inadequate or Imprecise Result or Readings	14/11/2025	14/11/2025
							F08	Hospitalization or Prolonged Hospitalization	A0908	Incorrect, Inadequate or Imprecise Result or Readings	14/11/2025	14/11/2025
							F15	Recognized Device or Procedural Complication	A0908	Incorrect, Inadequate or Imprecise Result or Readings	14/11/2025	14/11/2025
							F1905	Device Revision or Replacement	A0908	Incorrect, Inadequate or Imprecise Result or Readings	14/11/2025	14/11/2025
17812	2025	All other reportable incidents	47265	Dual-chamber implantable pacemaker, rate-responsive	CT1685	Cardiac pacemakers	E060106	Heart Block	A071102	Failure to Interrogate		
							F1905	Device Revision or Replacement	A071102	Failure to Interrogate		
					CT257	Implantable cardiac pacemakers	E060106	Heart Block	A071102	Failure to Interrogate		
							F1905	Device Revision or Replacement	A071102	Failure to Interrogate		
					CT2866	Implantable cardiac electrical stimulators	E060106	Heart Block	A071102	Failure to Interrogate		
							F1905	Device Revision or Replacement	A071102	Failure to Interrogate		
17813	2025	All other reportable incidents	60789	Intracardiac pacemaker			E2403	No Clinical Signs, Symptoms or Conditions	A0709	Device Sensing Problem		
									A090801	Signal Artifact/Noise		
							F26	No Health Consequences or Impact	A0709	Device Sensing Problem		
									A090801	Signal Artifact/Noise		
17814	2025	All other reportable incidents	47265	Dual-chamber implantable pacemaker, rate-responsive			E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
17815	2025	All other reportable incidents	47265	Dual-chamber implantable pacemaker, rate-responsive			E2403	No Clinical Signs, Symptoms or Conditions	A160105	False Alarm		
									A160503	Inappropriate or Unexpected Reset		
							F26	No Health Consequences or Impact	A160105	False Alarm		
									A160503	Inappropriate or Unexpected Reset		
17816	2025	All other reportable incidents	60789	Intracardiac pacemaker			E2403	No Clinical Signs, Symptoms or Conditions	A070102	High Capture Threshold	23/03/2026	23/03/2026
									A070104	Unstable Capture Threshold	23/03/2026	23/03/2026
									A070504	Premature Discharge of Battery	23/03/2026	23/03/2026
							F08	Hospitalization or Prolonged Hospitalization	A070102	High Capture Threshold	23/03/2026	23/03/2026
									A070104	Unstable Capture Threshold	23/03/2026	23/03/2026
									A070504	Premature Discharge of Battery	23/03/2026	23/03/2026
							F12	Serious Injury/ Illness/ Impairment	A070102	High Capture Threshold	23/03/2026	23/03/2026
									A070104	Unstable Capture Threshold	23/03/2026	23/03/2026
									A070504	Premature Discharge of Battery	23/03/2026	23/03/2026
							F1905	Device Revision or Replacement	A070102	High Capture Threshold	23/03/2026	23/03/2026
									A070104	Unstable Capture Threshold	23/03/2026	23/03/2026
									A070504	Premature Discharge of Battery	23/03/2026	23/03/2026
							F26	No Health Consequences or Impact	A070102	High Capture Threshold	23/03/2026	23/03/2026
									A070104	Unstable Capture Threshold	23/03/2026	23/03/2026
									A070504	Premature Discharge of Battery	23/03/2026	23/03/2026
17817	2025	All other reportable incidents	47265	Dual-chamber implantable pacemaker, rate-responsive			E1713	Pocket Erosion	A24	No Apparent Device Problem	12/09/2025	12/09/2025
							E1906	Unspecified Infection	A24	No Apparent Device Problem	12/09/2025	12/09/2025
							F0801	Intensive Care	A24	No Apparent Device Problem	12/09/2025	12/09/2025
							F1905	Device Revision or Replacement	A24	No Apparent Device Problem	12/09/2025	12/09/2025
17818	2025	All other reportable incidents	47265	Dual-chamber implantable pacemaker, rate-responsive			E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	07/01/2026	07/01/2026
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	07/01/2026	07/01/2026
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	07/01/2026	07/01/2026
17819	2025	All other reportable incidents	47265	Dual-chamber implantable pacemaker, rate-responsive			E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
17820	2025		60789	Intracardiac pacemaker			E2403	No Clinical Signs, Symptoms or Conditions	A25	No Apparent Adverse Event		
							F26	No Health Consequences or Impact	A25	No Apparent Adverse Event		

17821	2025	All other reportable incidents	47265	Dual-chamber implantable pacemaker, rate-responsive			E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
17822	2025	All other reportable incidents	47265	Dual-chamber implantable pacemaker, rate-responsive			E2403	No Clinical Signs, Symptoms or Conditions	A0908	Incorrect, Inadequate or Imprecise Result or Readings	12/11/2025	12/11/2025
							F08	Hospitalization or Prolonged Hospitalization	A0908	Incorrect, Inadequate or Imprecise Result or Readings	12/11/2025	12/11/2025
							F1905	Device Revision or Replacement	A0908	Incorrect, Inadequate or Imprecise Result or Readings	12/11/2025	12/11/2025
17823	2025	All other reportable incidents	47265	Dual-chamber implantable pacemaker, rate-responsive			E2403	No Clinical Signs, Symptoms or Conditions	A070909	Oversensing		
									A0712	Pacing Problem		
							F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A070909	Oversensing		
									A0712	Pacing Problem		
17824	2025	All other reportable incidents	47267	Single-chamber implantable pacemaker, rate-responsive			E2403	No Clinical Signs, Symptoms or Conditions	A071102	Failure to Interrogate		
							F26	No Health Consequences or Impact	A071102	Failure to Interrogate		
17825	2025	All other reportable incidents	47265	Dual-chamber implantable pacemaker, rate-responsive			E0618	Non Specific EKG/ECG Changes	A0712	Pacing Problem		
							E2403	No Clinical Signs, Symptoms or Conditions	A0712	Pacing Problem		
							F2307	Reprogramming of Device	A0712	Pacing Problem		
17826	2025	Unanticipated serious deterioration in state of health	47267	Single-chamber implantable pacemaker, rate-responsive			E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	27/11/2025	27/11/2025
									A22	Human-Device Interface Problem	27/11/2025	27/11/2025
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	27/11/2025	27/11/2025
									A22	Human-Device Interface Problem	27/11/2025	27/11/2025
							F12	Serious Injury/ Illness/ Impairment	A070504	Premature Discharge of Battery	27/11/2025	27/11/2025
									A22	Human-Device Interface Problem	27/11/2025	27/11/2025
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	27/11/2025	27/11/2025
									A22	Human-Device Interface Problem	27/11/2025	27/11/2025
17827	2025		47265	Dual-chamber implantable pacemaker, rate-responsive	CT2750	Implantable cardiac electrical stimulators and associated devices	E2401	Insufficient Information	A02	Manufacturing, Packaging or Shipping Problem		
									A070501	Battery Problem: High Impedance		
									A1106	Data Back-Up Problem		
							F06	Disruption of Subsequent Medical Procedure	A02	Manufacturing, Packaging or Shipping Problem		
									A070501	Battery Problem: High Impedance		
									A1106	Data Back-Up Problem		
17828	2025	All other reportable incidents	47265	Dual-chamber implantable pacemaker, rate-responsive	CT2750	Implantable cardiac electrical stimulators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A0201	Product Quality Problem		
									A070501	Battery Problem: High Impedance		
									A070504	Premature Discharge of Battery		
							F06	Disruption of Subsequent Medical Procedure	A0201	Product Quality Problem		
									A070501	Battery Problem: High Impedance		
									A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A0201	Product Quality Problem		
									A070501	Battery Problem: High Impedance		
									A070504	Premature Discharge of Battery		
17829	2025	All other reportable incidents	47265	Dual-chamber implantable pacemaker, rate-responsive			E2403	No Clinical Signs, Symptoms or Conditions	A070909	Oversensing		
									A1107	Data Problem		
							F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A070909	Oversensing		
									A1107	Data Problem		
17830	2025	All other reportable incidents	47265	Dual-chamber implantable pacemaker, rate-responsive			E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	01/12/2025	01/12/2025
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	01/12/2025	01/12/2025
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	01/12/2025	01/12/2025
17831	2025	All other reportable incidents	47265	Dual-chamber implantable pacemaker, rate-responsive			E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
17832	2025	All other reportable incidents	47265	Dual-chamber implantable pacemaker, rate-responsive			E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
17833	2025	All other reportable incidents	47267	Single-chamber implantable pacemaker, rate-responsive			E012207	Twitching	A070909	Oversensing	21/11/2025	21/11/2025
									A0712	Pacing Problem	21/11/2025	21/11/2025
									A0908	Incorrect, Inadequate or Imprecise Result or Readings	21/11/2025	21/11/2025
							E060104	Bradycardia	A070909	Oversensing	21/11/2025	21/11/2025
									A0712	Pacing Problem	21/11/2025	21/11/2025
									A0908	Incorrect, Inadequate or Imprecise Result or Readings	21/11/2025	21/11/2025
							F08	Hospitalization or Prolonged Hospitalization	A070909	Oversensing	21/11/2025	21/11/2025
									A0712	Pacing Problem	21/11/2025	21/11/2025
									A0908	Incorrect, Inadequate or Imprecise Result or Readings	21/11/2025	21/11/2025
							F1001	Absence of Treatment	A070909	Oversensing	21/11/2025	21/11/2025
									A0712	Pacing Problem	21/11/2025	21/11/2025
									A0908	Incorrect, Inadequate or Imprecise Result or Readings	21/11/2025	21/11/2025
							F1905	Device Revision or Replacement	A070909	Oversensing	21/11/2025	21/11/2025
									A0712	Pacing Problem	21/11/2025	21/11/2025
									A0908	Incorrect, Inadequate or Imprecise Result or Readings	21/11/2025	21/11/2025
17834	2025	All other reportable incidents	47265	Dual-chamber implantable pacemaker, rate-responsive			E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
17835	2025	All other reportable incidents	60789	Intracardiac pacemaker			E2403	No Clinical Signs, Symptoms or Conditions	A1107	Data Problem		
							F2307	Reprogramming of Device	A1107	Data Problem		
17836	2025	All other reportable incidents	60789	Intracardiac pacemaker			E2403	No Clinical Signs, Symptoms or Conditions	A1107	Data Problem		
							F2307	Reprogramming of Device	A1107	Data Problem		
17837	2025	All other reportable incidents	47265	Dual-chamber implantable pacemaker, rate-responsive			E2403	No Clinical Signs, Symptoms or Conditions	A0712	Pacing Problem		
							F26	No Health Consequences or Impact	A0712	Pacing Problem		
17838	2025	All other reportable incidents	47263	Cardiac resynchronization therapy implantable pacemaker	CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A0908	Incorrect, Inadequate or Imprecise Result or Readings		
							F26	No Health Consequences or Impact	A0908	Incorrect, Inadequate or Imprecise Result or Readings		
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A0908	Incorrect, Inadequate or Imprecise Result or Readings		
							F26	No Health Consequences or Impact	A0908	Incorrect, Inadequate or Imprecise Result or Readings		
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A0908	Incorrect, Inadequate or Imprecise Result or Readings		

							F26	No Health Consequences or Impact	A0908	Incorrect, Inadequate or Imprecise Result or Readings			
17839	2025		47263	Cardiac resynchronization therapy implantable pacemaker	CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A071205	Pacemaker Found in Back-Up Mode			
									A2205	Device Provides Uninformative Message			
									A27	Appropriate Device Problem Term/Code not Available			
							F2304	Prophylactic Treatment	A071205	Pacemaker Found in Back-Up Mode			
									A2205	Device Provides Uninformative Message			
									A27	Appropriate Device Problem Term/Code not Available			
							F26	No Health Consequences or Impact	A071205	Pacemaker Found in Back-Up Mode			
									A2205	Device Provides Uninformative Message			
									A27	Appropriate Device Problem Term/Code not Available			
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A071205	Pacemaker Found in Back-Up Mode			
									A2205	Device Provides Uninformative Message			
									A27	Appropriate Device Problem Term/Code not Available			
							F2304	Prophylactic Treatment	A071205	Pacemaker Found in Back-Up Mode			
									A2205	Device Provides Uninformative Message			
									A27	Appropriate Device Problem Term/Code not Available			
							F26	No Health Consequences or Impact	A071205	Pacemaker Found in Back-Up Mode			
									A2205	Device Provides Uninformative Message			
									A27	Appropriate Device Problem Term/Code not Available			
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A071205	Pacemaker Found in Back-Up Mode			
									A2205	Device Provides Uninformative Message			
									A27	Appropriate Device Problem Term/Code not Available			
							F2304	Prophylactic Treatment	A071205	Pacemaker Found in Back-Up Mode			
									A2205	Device Provides Uninformative Message			
									A27	Appropriate Device Problem Term/Code not Available			
							F26	No Health Consequences or Impact	A071205	Pacemaker Found in Back-Up Mode			
									A2205	Device Provides Uninformative Message			
									A27	Appropriate Device Problem Term/Code not Available			
17840	2025	All other reportable incidents	47265	Dual-chamber implantable pacemaker, rate-responsive			E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery			
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery			
17841	2025	All other reportable incidents	47265	Dual-chamber implantable pacemaker, rate-responsive			E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery			
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery			
17842	2025		35852	Single-chamber implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E0102	Brain Injury	A0105	Lack of Effect			
									A0203	Defective Device			
									A0713	Defibrillation/Stimulation Problem			
									A0721	Circuit Failure			
							F02	Death	A0105	Lack of Effect			
									A0203	Defective Device			
									A0713	Defibrillation/Stimulation Problem			
									A0721	Circuit Failure			
17843	2025	All other reportable incidents	47265	Dual-chamber implantable pacemaker, rate-responsive			E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery			
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery			
17844	2025	All other reportable incidents	47265	Dual-chamber implantable pacemaker, rate-responsive			E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	01/12/2025	01/12/2025	
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	01/12/2025	01/12/2025	
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	01/12/2025	01/12/2025	
17845	2025	All other reportable incidents	47265	Dual-chamber implantable pacemaker, rate-responsive			E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery			
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery			
17846	2025	All other reportable incidents	47265	Dual-chamber implantable pacemaker, rate-responsive			E2403	No Clinical Signs, Symptoms or Conditions	A0908	Incorrect, Inadequate or Imprecise Result or Readings	24/11/2025	24/11/2025	
							F08	Hospitalization or Prolonged Hospitalization	A0908	Incorrect, Inadequate or Imprecise Result or Readings	24/11/2025	24/11/2025	
							F1905	Device Revision or Replacement	A0908	Incorrect, Inadequate or Imprecise Result or Readings	24/11/2025	24/11/2025	
17847	2025	All other reportable incidents	47265	Dual-chamber implantable pacemaker, rate-responsive			E2403	No Clinical Signs, Symptoms or Conditions	A05	Mechanical Problem	23/01/2025	23/01/2025	
							F26	No Health Consequences or Impact	A05	Mechanical Problem	23/01/2025	23/01/2025	
17848	2025		47265	Dual-chamber implantable pacemaker, rate-responsive	CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A070501	Battery Problem: High Impedance			
							F26	No Health Consequences or Impact	A070501	Battery Problem: High Impedance			
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A070501	Battery Problem: High Impedance			
							F26	No Health Consequences or Impact	A070501	Battery Problem: High Impedance			
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A070501	Battery Problem: High Impedance			
							F26	No Health Consequences or Impact	A070501	Battery Problem: High Impedance			
17849	2025	All other reportable incidents	60789	Intracardiac pacemaker			E2403	No Clinical Signs, Symptoms or Conditions	A1107	Data Problem			
							F26	No Health Consequences or Impact	A1107	Data Problem			
17850	2025	All other reportable incidents	60789	Intracardiac pacemaker			E0619	Pericardial Effusion	A24	No Apparent Device Problem			
							E2311	Discomfort	A24	No Apparent Device Problem			
							F12	Serious Injury/ Illness/ Impairment	A24	No Apparent Device Problem			
17851	2025	All other reportable incidents	47267	Single-chamber implantable pacemaker, rate-responsive			E2403	No Clinical Signs, Symptoms or Conditions	A0908	Incorrect, Inadequate or Imprecise Result or Readings			
							F26	No Health Consequences or Impact	A0908	Incorrect, Inadequate or Imprecise Result or Readings			
17852	2025	All other reportable incidents	47265	Dual-chamber implantable pacemaker, rate-responsive			E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery			
									A072201	High Impedance			
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery			
									A072201	High Impedance			
17853	2025	All other reportable incidents	47265	Dual-chamber implantable pacemaker, rate-responsive			E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	05/01/2026	05/01/2026	
							F1901	Additional Surgery	A070504	Premature Discharge of Battery	05/01/2026	05/01/2026	
17854	2025	Death	47265	Dual-chamber implantable pacemaker, rate-responsive			E2401	Insufficient Information	A070504	Premature Discharge of Battery			
							F02	Death	A070504	Premature Discharge of Battery			
17855	2025	All other reportable incidents	47267	Single-chamber implantable pacemaker, rate-responsive			E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem			
									A070504	Premature Discharge of Battery			
							F26	No Health Consequences or Impact	A0705	Battery Problem			

									A070504	Premature Discharge of Battery		
17856	2025	All other reportable incidents	47265	Dual-chamber implantable pacemaker, rate-responsive			E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
17857	2025	All other reportable incidents	47265	Dual-chamber implantable pacemaker, rate-responsive			E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
17858	2025	All other reportable incidents	60789	Intracardiac pacemaker			E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
17859	2025	All other reportable incidents	60789	Intracardiac pacemaker			E2403	No Clinical Signs, Symptoms or Conditions	A1107	Data Problem		
							F26	No Health Consequences or Impact	A1107	Data Problem		
17860	2025	All other reportable incidents	60789	Intracardiac pacemaker			E2403	No Clinical Signs, Symptoms or Conditions	A1107	Data Problem		
							F26	No Health Consequences or Impact	A1107	Data Problem		
17861	2025		18145	Cardiac myoplasty electrical stimulation system	CT1684	Cardiac pacemakers and associated devices	E060110	Ventricular Fibrillation	A0722	Impedance Problem		
							F02	Death	A0722	Impedance Problem		
					CT2866	Implantable cardiac electrical stimulators	E060110	Ventricular Fibrillation	A0722	Impedance Problem		
							F02	Death	A0722	Impedance Problem		
17862	2025	All other reportable incidents	47267	Single-chamber implantable pacemaker, rate-responsive			E2403	No Clinical Signs, Symptoms or Conditions	A070909	Oversensing	24/11/2025	24/11/2025
									A0712	Pacing Problem	24/11/2025	24/11/2025
									A0908	Incorrect, Inadequate or Imprecise Result or Readings	24/11/2025	24/11/2025
							F08	Hospitalization or Prolonged Hospitalization	A070909	Oversensing	24/11/2025	24/11/2025
									A0712	Pacing Problem	24/11/2025	24/11/2025
									A0908	Incorrect, Inadequate or Imprecise Result or Readings	24/11/2025	24/11/2025
							F1001	Absence of Treatment	A070909	Oversensing	24/11/2025	24/11/2025
									A0712	Pacing Problem	24/11/2025	24/11/2025
									A0908	Incorrect, Inadequate or Imprecise Result or Readings	24/11/2025	24/11/2025
							F1905	Device Revision or Replacement	A070909	Oversensing	24/11/2025	24/11/2025
									A0712	Pacing Problem	24/11/2025	24/11/2025
									A0908	Incorrect, Inadequate or Imprecise Result or Readings	24/11/2025	24/11/2025
17863	2025	All other reportable incidents	47265	Dual-chamber implantable pacemaker, rate-responsive			E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
17864	2025	All other reportable incidents	47265	Dual-chamber implantable pacemaker, rate-responsive			E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem	22/04/2025	22/04/2025
							F08	Hospitalization or Prolonged Hospitalization	A0705	Battery Problem	22/04/2025	22/04/2025
							F1905	Device Revision or Replacement	A0705	Battery Problem	22/04/2025	22/04/2025
17865	2025	All other reportable incidents	47265	Dual-chamber implantable pacemaker, rate-responsive			E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	08/01/2026	08/01/2026
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	08/01/2026	08/01/2026
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	08/01/2026	08/01/2026
17866	2025	Unanticipated serious deterioration in state of health	47265	Dual-chamber implantable pacemaker, rate-responsive			E012207	Twitching	A0712	Pacing Problem	02/12/2025	02/12/2025
							E1705	Burning Sensation	A0712	Pacing Problem	02/12/2025	02/12/2025
							F08	Hospitalization or Prolonged Hospitalization	A0712	Pacing Problem	02/12/2025	02/12/2025
							F12	Serious Injury/ Illness/ Impairment	A0712	Pacing Problem	02/12/2025	02/12/2025
							F1903	Device Explantation	A0712	Pacing Problem	02/12/2025	02/12/2025
							F2307	Reprogramming of Device	A0712	Pacing Problem	02/12/2025	02/12/2025
17867	2025	All other reportable incidents	47265	Dual-chamber implantable pacemaker, rate-responsive			E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
17868	2025	All other reportable incidents	47265	Dual-chamber implantable pacemaker, rate-responsive			E2403	No Clinical Signs, Symptoms or Conditions	A070909	Oversensing		
									A0712	Pacing Problem		
									A090801	Signal Artifact/Noise		
							F1001	Absence of Treatment	A070909	Oversensing		
									A0712	Pacing Problem		
									A090801	Signal Artifact/Noise		
17869	2025	All other reportable incidents	47265	Dual-chamber implantable pacemaker, rate-responsive			E2403	No Clinical Signs, Symptoms or Conditions	A070101	Failure to Capture		
									A070910	Undersensing		
									A0712	Pacing Problem		
							F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A070101	Failure to Capture		
									A070910	Undersensing		
									A0712	Pacing Problem		
17870	2025	All other reportable incidents	60789	Intracardiac pacemaker			E233001	Chest Pain	A0701	Capturing Problem		
									A0722	Impedance Problem		
							F11	Non-Serious Injury/ Illness/ Impairment	A0701	Capturing Problem		
									A0722	Impedance Problem		
17871	2025	All other reportable incidents	47265	Dual-chamber implantable pacemaker, rate-responsive			E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
17872	2025	All other reportable incidents	47265	Dual-chamber implantable pacemaker, rate-responsive			E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	30/10/2025	30/10/2025
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	30/10/2025	30/10/2025
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	30/10/2025	30/10/2025
17873	2025	All other reportable incidents	47265	Dual-chamber implantable pacemaker, rate-responsive	CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A0908	Incorrect, Inadequate or Imprecise Result or Readings	28/11/2025	28/11/2025
									A1107	Data Problem	28/11/2025	28/11/2025
							F08	Hospitalization or Prolonged Hospitalization	A0908	Incorrect, Inadequate or Imprecise Result or Readings	28/11/2025	28/11/2025
									A1107	Data Problem	28/11/2025	28/11/2025
							F1905	Device Revision or Replacement	A0908	Incorrect, Inadequate or Imprecise Result or Readings	28/11/2025	28/11/2025
									A1107	Data Problem	28/11/2025	28/11/2025
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A0908	Incorrect, Inadequate or Imprecise Result or Readings	28/11/2025	28/11/2025
									A1107	Data Problem	28/11/2025	28/11/2025
							F08	Hospitalization or Prolonged Hospitalization	A0908	Incorrect, Inadequate or Imprecise Result or Readings	28/11/2025	28/11/2025
									A1107	Data Problem	28/11/2025	28/11/2025
							F1905	Device Revision or Replacement	A0908	Incorrect, Inadequate or Imprecise Result or Readings	28/11/2025	28/11/2025
									A1107	Data Problem	28/11/2025	28/11/2025
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A0908	Incorrect, Inadequate or Imprecise Result or Readings	28/11/2025	28/11/2025
									A1107	Data Problem	28/11/2025	28/11/2025
							F08	Hospitalization or Prolonged Hospitalization	A0908	Incorrect, Inadequate or Imprecise Result or Readings	28/11/2025	28/11/2025

									A1107	Data Problem	28/11/2025	28/11/2025
							F1905	Device Revision or Replacement	A0908	Incorrect, Inadequate or Imprecise Result or Readings	28/11/2025	28/11/2025
									A1107	Data Problem	28/11/2025	28/11/2025
17874	2025	All other reportable incidents	47265	Dual-chamber implantable pacemaker, rate-responsive			E2403	No Clinical Signs, Symptoms or Conditions	A0711	Interrogation Problem		
							F26	No Health Consequences or Impact	A0711	Interrogation Problem		
17875	2025	All other reportable incidents	47265	Dual-chamber implantable pacemaker, rate-responsive			E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
17876	2025	All other reportable incidents	60789	Intracardiac pacemaker			E2403	No Clinical Signs, Symptoms or Conditions	A1107	Data Problem		
							F26	No Health Consequences or Impact	A1107	Data Problem		
17877	2025	All other reportable incidents	47265	Dual-chamber implantable pacemaker, rate-responsive			E2403	No Clinical Signs, Symptoms or Conditions	A070101	Failure to Capture		
							F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A070101	Failure to Capture		
17878	2025	All other reportable incidents	47265	Dual-chamber implantable pacemaker, rate-responsive			E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
									A0722	Impedance Problem		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
									A0722	Impedance Problem		
17879	2025	All other reportable incidents	60789	Intracardiac pacemaker			E2403	No Clinical Signs, Symptoms or Conditions	A1107	Data Problem		
							F26	No Health Consequences or Impact	A1107	Data Problem		
17880	2025	All other reportable incidents	60789	Intracardiac pacemaker			E2403	No Clinical Signs, Symptoms or Conditions	A1107	Data Problem		
							F26	No Health Consequences or Impact	A1107	Data Problem		
17881	2025	All other reportable incidents	60789	Intracardiac pacemaker			E2403	No Clinical Signs, Symptoms or Conditions	A1107	Data Problem		
							F26	No Health Consequences or Impact	A1107	Data Problem		
17882	2025	All other reportable incidents	60789	Intracardiac pacemaker			E2403	No Clinical Signs, Symptoms or Conditions	A1107	Data Problem		
							F27	Problem Identified During Non-Clinical Procedure	A1107	Data Problem		
17883	2025	All other reportable incidents	60789	Intracardiac pacemaker			E2403	No Clinical Signs, Symptoms or Conditions	A1107	Data Problem		
							F26	No Health Consequences or Impact	A1107	Data Problem		
17884	2025	All other reportable incidents	47265	Dual-chamber implantable pacemaker, rate-responsive			E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
17885	2025	All other reportable incidents	60789	Intracardiac pacemaker			E2403	No Clinical Signs, Symptoms or Conditions	A1107	Data Problem		
							F26	No Health Consequences or Impact	A1107	Data Problem		
17886	2025		47267	Single-chamber implantable pacemaker, rate-responsive			E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	27/11/2025	27/11/2025
							F1901	Additional Surgery	A24	No Apparent Device Problem	27/11/2025	27/11/2025
17887	2025	All other reportable incidents	47265	Dual-chamber implantable pacemaker, rate-responsive			E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
17888	2025	All other reportable incidents	47265	Dual-chamber implantable pacemaker, rate-responsive			E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
17889	2025	All other reportable incidents	47265	Dual-chamber implantable pacemaker, rate-responsive			E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
17890	2025	All other reportable incidents	47265	Dual-chamber implantable pacemaker, rate-responsive	CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	01/03/2026	01/03/2026
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	01/03/2026	01/03/2026
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	01/03/2026	01/03/2026
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	01/03/2026	01/03/2026
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	01/03/2026	01/03/2026
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	01/03/2026	01/03/2026
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	01/03/2026	01/03/2026
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	01/03/2026	01/03/2026
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	01/03/2026	01/03/2026
17891	2025	All other reportable incidents	47265	Dual-chamber implantable pacemaker, rate-responsive			E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	03/12/2025	03/12/2025
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	03/12/2025	03/12/2025
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	03/12/2025	03/12/2025
17892	2025	All other reportable incidents	47265	Dual-chamber implantable pacemaker, rate-responsive			E2007	Fall	A0711	Interrogation Problem	09/01/2026	09/01/2026
							F08	Hospitalization or Prolonged Hospitalization	A0711	Interrogation Problem	09/01/2026	09/01/2026
							F1905	Device Revision or Replacement	A0711	Interrogation Problem	09/01/2026	09/01/2026
							F26	No Health Consequences or Impact	A0711	Interrogation Problem	09/01/2026	09/01/2026
17893	2025	All other reportable incidents	47267	Single-chamber implantable pacemaker, rate-responsive			E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
17894	2025	All other reportable incidents	47267	Single-chamber implantable pacemaker, rate-responsive			E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
17895	2025	All other reportable incidents	63188	Left-ventricular wireless pacing system sensor/transmitter			E0623	Unspecified Heart Problem	A0708	Power Problem		
							F26	No Health Consequences or Impact	A0708	Power Problem		
17896	2025	All other reportable incidents	47265	Dual-chamber implantable pacemaker, rate-responsive			E0717	Dyspnea	A23	Use of Device Problem		
							E1901	Bacterial Infection	A23	Use of Device Problem		
							E2401	Insufficient Information	A23	Use of Device Problem		
							E2403	No Clinical Signs, Symptoms or Conditions	A23	Use of Device Problem		
							F12	Serious Injury/ Illness/ Impairment	A23	Use of Device Problem		
							F2303	Medication Required	A23	Use of Device Problem		
							F2307	Reprogramming of Device	A23	Use of Device Problem		
							F24	Insufficient Health Impact Information	A23	Use of Device Problem		
17897	2025	All other reportable incidents	47265	Dual-chamber implantable pacemaker, rate-responsive			E2403	No Clinical Signs, Symptoms or Conditions	A05	Mechanical Problem		
							F26	No Health Consequences or Impact	A05	Mechanical Problem		
17898	2025	All other reportable incidents	47265	Dual-chamber implantable pacemaker, rate-responsive			E2403	No Clinical Signs, Symptoms or Conditions	A0908	Incorrect, Inadequate or Imprecise Result or Readings	04/12/2025	04/12/2025
							F08	Hospitalization or Prolonged Hospitalization	A0908	Incorrect, Inadequate or Imprecise Result or Readings	04/12/2025	04/12/2025
							F1905	Device Revision or Replacement	A0908	Incorrect, Inadequate or Imprecise Result or Readings	04/12/2025	04/12/2025
17899	2025	All other reportable incidents	47265	Dual-chamber implantable pacemaker, rate-responsive			E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem		
							F26	No Health Consequences or Impact	A0705	Battery Problem		
17900	2025		47265	Dual-chamber implantable pacemaker, rate-responsive	CT1685	Cardiac pacemakers	E012207	Twitching	A0711	Interrogation Problem		
									A071205	Pacemaker Found in Back-Up Mode		
							F2307	Reprogramming of Device	A0711	Interrogation Problem		
									A071205	Pacemaker Found in Back-Up Mode		
					CT257	Implantable cardiac pacemakers	E012207	Twitching	A0711	Interrogation Problem		
									A071205	Pacemaker Found in Back-Up Mode		
							F2307	Reprogramming of Device	A0711	Interrogation Problem		
									A071205	Pacemaker Found in Back-Up Mode		
					CT2866	Implantable cardiac electrical stimulators	E012207	Twitching	A0711	Interrogation Problem		
									A071205	Pacemaker Found in Back-Up Mode		
							F2307	Reprogramming of Device	A0711	Interrogation Problem		

									A071205	Pacemaker Found in Back-Up Mode		
17901	2025	All other reportable incidents	47265	Dual-chamber implantable pacemaker, rate-responsive			E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
									A0711	Interrogation Problem		
									A1107	Data Problem		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
									A0711	Interrogation Problem		
									A1107	Data Problem		
17902	2025	All other reportable incidents	47267	Single-chamber implantable pacemaker, rate-responsive			E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	13/11/2025	13/11/2025
							F1901	Additional Surgery	A24	No Apparent Device Problem	13/11/2025	13/11/2025
17903	2025	All other reportable incidents	47265	Dual-chamber implantable pacemaker, rate-responsive			E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
17904	2025	All other reportable incidents	47265	Dual-chamber implantable pacemaker, rate-responsive			E011903	Syncope/Fainting	A070504	Premature Discharge of Battery	10/12/2025	10/12/2025
									A071102	Failure to Interrogate	10/12/2025	10/12/2025
									A071204	No Pacing	10/12/2025	10/12/2025
							F1901	Additional Surgery	A070504	Premature Discharge of Battery	10/12/2025	10/12/2025
									A071102	Failure to Interrogate	10/12/2025	10/12/2025
									A071204	No Pacing	10/12/2025	10/12/2025
17905	2025	All other reportable incidents	47265	Dual-chamber implantable pacemaker, rate-responsive			E2403	No Clinical Signs, Symptoms or Conditions	A070909	Oversensing		
									A090801	Signal Artifact/Noise		
							F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A070909	Oversensing		
									A090801	Signal Artifact/Noise		
							F2203	Imaging Required	A070909	Oversensing		
									A090801	Signal Artifact/Noise		
17906	2025	All other reportable incidents	47265	Dual-chamber implantable pacemaker, rate-responsive			E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
17907	2025		47265	Dual-chamber implantable pacemaker, rate-responsive	CT1685	Cardiac pacemakers	E0118	Intracranial Hemorrhage	A071205	Pacemaker Found in Back-Up Mode		
							F11	Non-Serious Injury/ Illness/ Impairment	A071205	Pacemaker Found in Back-Up Mode		
					CT257	Implantable cardiac pacemakers	E0118	Intracranial Hemorrhage	A071205	Pacemaker Found in Back-Up Mode		
							F11	Non-Serious Injury/ Illness/ Impairment	A071205	Pacemaker Found in Back-Up Mode		
					CT2866	Implantable cardiac electrical stimulators	E0118	Intracranial Hemorrhage	A071205	Pacemaker Found in Back-Up Mode		
							F11	Non-Serious Injury/ Illness/ Impairment	A071205	Pacemaker Found in Back-Up Mode		
17908	2025		35852	Single-chamber implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F1903	Device Explantation	A070504	Premature Discharge of Battery		
17909	2025		47265	Dual-chamber implantable pacemaker, rate-responsive	CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A0201	Product Quality Problem		
							F26	No Health Consequences or Impact	A0201	Product Quality Problem		
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A0201	Product Quality Problem		
							F26	No Health Consequences or Impact	A0201	Product Quality Problem		
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A0201	Product Quality Problem		
							F26	No Health Consequences or Impact	A0201	Product Quality Problem		
17910	2025	All other reportable incidents	47263	Cardiac resynchronization therapy implantable pacemaker	CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A070501	Battery Problem: High Impedance		
									A070504	Premature Discharge of Battery		
							F1905	Device Revision or Replacement	A070501	Battery Problem: High Impedance		
									A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070501	Battery Problem: High Impedance		
									A070504	Premature Discharge of Battery		
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A070501	Battery Problem: High Impedance		
									A070504	Premature Discharge of Battery		
							F1905	Device Revision or Replacement	A070501	Battery Problem: High Impedance		
									A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070501	Battery Problem: High Impedance		
									A070504	Premature Discharge of Battery		
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A070501	Battery Problem: High Impedance		
									A070504	Premature Discharge of Battery		
							F1905	Device Revision or Replacement	A070501	Battery Problem: High Impedance		
									A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070501	Battery Problem: High Impedance		
									A070504	Premature Discharge of Battery		
17911	2025	All other reportable incidents	47265	Dual-chamber implantable pacemaker, rate-responsive			E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	28/01/2026	28/01/2026
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	28/01/2026	28/01/2026
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	28/01/2026	28/01/2026
17912	2025	All other reportable incidents	47265	Dual-chamber implantable pacemaker, rate-responsive			E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	26/01/2026	26/01/2026
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	26/01/2026	26/01/2026
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	26/01/2026	26/01/2026
17913	2025	All other reportable incidents	47265	Dual-chamber implantable pacemaker, rate-responsive			E2403	No Clinical Signs, Symptoms or Conditions	A0712	Pacing Problem		
									A0908	Incorrect, Inadequate or Imprecise Result or Readings		
							F26	No Health Consequences or Impact	A0712	Pacing Problem		
									A0908	Incorrect, Inadequate or Imprecise Result or Readings		
17914	2025	All other reportable incidents	47265	Dual-chamber implantable pacemaker, rate-responsive	CT2750	Implantable cardiac electrical stimulators and associated devices	E0112	Dizziness	A0202	Defective Component	16/10/2025	16/10/2025
									A070501	Battery Problem: High Impedance	16/10/2025	16/10/2025
									A072201	High Impedance	16/10/2025	16/10/2025
									A0908	Incorrect, Inadequate or Imprecise Result or Readings	16/10/2025	16/10/2025
									A160503	Inappropriate or Unexpected Reset	16/10/2025	16/10/2025
							E2403	No Clinical Signs, Symptoms or Conditions	A0202	Defective Component	16/10/2025	16/10/2025
									A070501	Battery Problem: High Impedance	16/10/2025	16/10/2025
									A072201	High Impedance	16/10/2025	16/10/2025
									A0908	Incorrect, Inadequate or Imprecise Result or Readings	16/10/2025	16/10/2025
									A160503	Inappropriate or Unexpected Reset	16/10/2025	16/10/2025
							F08	Hospitalization or Prolonged Hospitalization	A0202	Defective Component	16/10/2025	16/10/2025
									A070501	Battery Problem: High Impedance	16/10/2025	16/10/2025
									A072201	High Impedance	16/10/2025	16/10/2025
									A0908	Incorrect, Inadequate or Imprecise Result or Readings	16/10/2025	16/10/2025
									A160503	Inappropriate or Unexpected Reset	16/10/2025	16/10/2025
							F15	Recognized Device or Procedural Complication	A0202	Defective Component	16/10/2025	16/10/2025
									A070501	Battery Problem: High Impedance	16/10/2025	16/10/2025

									A072201	High Impedance	16/10/2025	16/10/2025
									A0908	Incorrect, Inadequate or Imprecise Result or Readings	16/10/2025	16/10/2025
									A160503	Inappropriate or Unexpected Reset	16/10/2025	16/10/2025
							F1905	Device Revision or Replacement	A0202	Defective Component	16/10/2025	16/10/2025
									A070501	Battery Problem: High Impedance	16/10/2025	16/10/2025
									A072201	High Impedance	16/10/2025	16/10/2025
									A0908	Incorrect, Inadequate or Imprecise Result or Readings	16/10/2025	16/10/2025
									A160503	Inappropriate or Unexpected Reset	16/10/2025	16/10/2025
17915	2025	All other reportable incidents	47265	Dual-chamber implantable pacemaker, rate-responsive			E2403	No Clinical Signs, Symptoms or Conditions	A070910	Undersensing		
							F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A070910	Undersensing		
17916	2025	All other reportable incidents	47265	Dual-chamber implantable pacemaker, rate-responsive			E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
							F23	Unexpected Medical Intervention	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
17917	2025	All other reportable incidents	47267	Single-chamber implantable pacemaker, rate-responsive	CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions				
							F08	Hospitalization or Prolonged Hospitalization				
							F1905	Device Revision or Replacement				
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions				
							F08	Hospitalization or Prolonged Hospitalization				
							F1905	Device Revision or Replacement				
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions				
							F08	Hospitalization or Prolonged Hospitalization				
							F1905	Device Revision or Replacement				
17918	2025	All other reportable incidents	47265	Dual-chamber implantable pacemaker, rate-responsive			E2403	No Clinical Signs, Symptoms or Conditions	A0908	Incorrect, Inadequate or Imprecise Result or Readings	10/12/2025	10/12/2025
							F08	Hospitalization or Prolonged Hospitalization	A0908	Incorrect, Inadequate or Imprecise Result or Readings	10/12/2025	10/12/2025
							F1905	Device Revision or Replacement	A0908	Incorrect, Inadequate or Imprecise Result or Readings	10/12/2025	10/12/2025
17919	2025		47265	Dual-chamber implantable pacemaker, rate-responsive	CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A071205	Pacemaker Found in Back-Up Mode		
									A27	Appropriate Device Problem Term/Code not Available		
							F28	Appropriate Health Impact Term/Code not Available	A071205	Pacemaker Found in Back-Up Mode		
									A27	Appropriate Device Problem Term/Code not Available		
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A071205	Pacemaker Found in Back-Up Mode		
									A27	Appropriate Device Problem Term/Code not Available		
							F28	Appropriate Health Impact Term/Code not Available	A071205	Pacemaker Found in Back-Up Mode		
									A27	Appropriate Device Problem Term/Code not Available		
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A071205	Pacemaker Found in Back-Up Mode		
									A27	Appropriate Device Problem Term/Code not Available		
							F28	Appropriate Health Impact Term/Code not Available	A071205	Pacemaker Found in Back-Up Mode		
									A27	Appropriate Device Problem Term/Code not Available		
17920	2025		47265	Dual-chamber implantable pacemaker, rate-responsive			E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	18/12/2025	18/12/2025
							F1901	Additional Surgery	A24	No Apparent Device Problem	18/12/2025	18/12/2025
17921	2025	All other reportable incidents	60789	Intracardiac pacemaker			E060102	Atrial Fibrillation	A0709	Device Sensing Problem		
									A1107	Data Problem		
							E2403	No Clinical Signs, Symptoms or Conditions	A0709	Device Sensing Problem		
									A1107	Data Problem		
							F12	Serious Injury/ Illness/ Impairment	A0709	Device Sensing Problem		
									A1107	Data Problem		
							F2307	Reprogramming of Device	A0709	Device Sensing Problem		
									A1107	Data Problem		
17922	2025	All other reportable incidents	47265	Dual-chamber implantable pacemaker, rate-responsive			E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	28/01/2026	28/01/2026
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	28/01/2026	28/01/2026
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	28/01/2026	28/01/2026
17923	2025	All other reportable incidents	47265	Dual-chamber implantable pacemaker, rate-responsive			E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	08/03/2026	08/03/2026
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	08/03/2026	08/03/2026
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	08/03/2026	08/03/2026
17924	2025	All other reportable incidents	47265	Dual-chamber implantable pacemaker, rate-responsive	CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	05/12/2025	05/12/2025
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	05/12/2025	05/12/2025
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	05/12/2025	05/12/2025
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	05/12/2025	05/12/2025
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	05/12/2025	05/12/2025
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	05/12/2025	05/12/2025
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	05/12/2025	05/12/2025
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	05/12/2025	05/12/2025
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	05/12/2025	05/12/2025
17925	2025	All other reportable incidents	47265	Dual-chamber implantable pacemaker, rate-responsive			E2403	No Clinical Signs, Symptoms or Conditions	A070909	Oversensing		
									A070910	Undersensing		
							F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A070909	Oversensing		
									A070910	Undersensing		
17926	2025	All other reportable incidents	63188	Left-ventricular wireless pacing system sensor/transmitter			E0623	Unspecified Heart Problem	A25	No Apparent Adverse Event		
							F26	No Health Consequences or Impact	A25	No Apparent Adverse Event		
17927	2025	All other reportable incidents	60789	Intracardiac pacemaker			E2403	No Clinical Signs, Symptoms or Conditions	A0504	Leak/Splash		
									A24	No Apparent Device Problem		
							F26	No Health Consequences or Impact	A0504	Leak/Splash		
									A24	No Apparent Device Problem		
17928	2025	All other reportable incidents	35852	Single-chamber implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A160601	Premature Elective Replacement Indicator		
							F1903	Device Explantation	A160601	Premature Elective Replacement Indicator		
17929	2025	All other reportable incidents	47265	Dual-chamber implantable pacemaker, rate-responsive			E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	07/01/2026	07/01/2026
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	07/01/2026	07/01/2026
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	07/01/2026	07/01/2026
17930	2025	All other reportable incidents	47265	Dual-chamber implantable pacemaker, rate-			E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		

				responsive			F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
17931	2025	All other reportable incidents	47265	Dual-chamber implantable pacemaker, rate-responsive			E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery		
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery		
17932	2025	All other reportable incidents	47265	Dual-chamber implantable pacemaker, rate-responsive			E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
17933	2025	All other reportable incidents	47265	Dual-chamber implantable pacemaker, rate-responsive			E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
17934	2025	All other reportable incidents	47265	Dual-chamber implantable pacemaker, rate-responsive			E2403	No Clinical Signs, Symptoms or Conditions	A070909	Oversensing		
									A072201	High Impedance		
									A090801	Signal Artifact/Noise		
							F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A070909	Oversensing		
									A072201	High Impedance		
									A090801	Signal Artifact/Noise		
17935	2025	All other reportable incidents	47265	Dual-chamber implantable pacemaker, rate-responsive			E2403	No Clinical Signs, Symptoms or Conditions	A0712	Pacing Problem		
									A23	Use of Device Problem		
							F26	No Health Consequences or Impact	A0712	Pacing Problem		
									A23	Use of Device Problem		
17936	2025	All other reportable incidents	47267	Single-chamber implantable pacemaker, rate-responsive			E2403	No Clinical Signs, Symptoms or Conditions	A072202	Low Impedance		
							F26	No Health Consequences or Impact	A072202	Low Impedance		
17937	2025	All other reportable incidents	47265	Dual-chamber implantable pacemaker, rate-responsive			E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
17938	2025	All other reportable incidents	47267	Single-chamber implantable pacemaker, rate-responsive			E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
17939	2025	All other reportable incidents	47265	Dual-chamber implantable pacemaker, rate-responsive			E0601	Arrhythmia	A071102	Failure to Interrogate	16/12/2025	16/12/2025
									A071204	No Pacing	16/12/2025	16/12/2025
							E0717	Dyspnea	A071102	Failure to Interrogate	16/12/2025	16/12/2025
									A071204	No Pacing	16/12/2025	16/12/2025
							E2312	Fatigue	A071102	Failure to Interrogate	16/12/2025	16/12/2025
									A071204	No Pacing	16/12/2025	16/12/2025
							F1901	Additional Surgery	A071102	Failure to Interrogate	16/12/2025	16/12/2025
									A071204	No Pacing	16/12/2025	16/12/2025
17940	2025	All other reportable incidents	47265	Dual-chamber implantable pacemaker, rate-responsive			E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	17/12/2025	17/12/2025
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	17/12/2025	17/12/2025
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	17/12/2025	17/12/2025
17941	2025	All other reportable incidents	47265	Dual-chamber implantable pacemaker, rate-responsive			E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
17942	2025	All other reportable incidents	47265	Dual-chamber implantable pacemaker, rate-responsive			E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
17943	2025	All other reportable incidents	47265	Dual-chamber implantable pacemaker, rate-responsive			E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	15/12/2025	15/12/2025
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	15/12/2025	15/12/2025
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	15/12/2025	15/12/2025
17944	2025		47263	Cardiac resynchronization therapy implantable pacemaker	CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A071205	Pacemaker Found in Back-Up Mode		
									A27	Appropriate Device Problem Term/Code not Available		
							F26	No Health Consequences or Impact	A071205	Pacemaker Found in Back-Up Mode		
									A27	Appropriate Device Problem Term/Code not Available		
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A071205	Pacemaker Found in Back-Up Mode		
									A27	Appropriate Device Problem Term/Code not Available		
							F26	No Health Consequences or Impact	A071205	Pacemaker Found in Back-Up Mode		
									A27	Appropriate Device Problem Term/Code not Available		
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A071205	Pacemaker Found in Back-Up Mode		
									A27	Appropriate Device Problem Term/Code not Available		
							F26	No Health Consequences or Impact	A071205	Pacemaker Found in Back-Up Mode		
									A27	Appropriate Device Problem Term/Code not Available		
17945	2025	All other reportable incidents	47267	Single-chamber implantable pacemaker, rate-responsive			E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
17946	2025		47265	Dual-chamber implantable pacemaker, rate-responsive	CT1685	Cardiac pacemakers	E0112	Dizziness	A0705	Battery Problem		
							E2007	Fall	A0705	Battery Problem		
							F11	Non-Serious Injury/ Illness/ Impairment	A0705	Battery Problem		
					CT257	Implantable cardiac pacemakers	E0112	Dizziness	A0705	Battery Problem		
							E2007	Fall	A0705	Battery Problem		
							F11	Non-Serious Injury/ Illness/ Impairment	A0705	Battery Problem		
					CT2866	Implantable cardiac electrical stimulators	E0112	Dizziness	A0705	Battery Problem		
							E2007	Fall	A0705	Battery Problem		
							F11	Non-Serious Injury/ Illness/ Impairment	A0705	Battery Problem		
17947	2025	All other reportable incidents	47265	Dual-chamber implantable pacemaker, rate-responsive	CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A0105	Lack of Effect		
									A070504	Premature Discharge of Battery		
									A070907	Loss of Threshold		
							F26	No Health Consequences or Impact	A0105	Lack of Effect		
									A070504	Premature Discharge of Battery		
									A070907	Loss of Threshold		
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A0105	Lack of Effect		
									A070504	Premature Discharge of Battery		
									A070907	Loss of Threshold		
							F26	No Health Consequences or Impact	A0105	Lack of Effect		
									A070504	Premature Discharge of Battery		
									A070907	Loss of Threshold		
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A0105	Lack of Effect		
									A070504	Premature Discharge of Battery		
									A070907	Loss of Threshold		
							F26	No Health Consequences or Impact	A0105	Lack of Effect		
									A070504	Premature Discharge of Battery		
									A070907	Loss of Threshold		

17948	2025	All other reportable incidents	47265	Dual-chamber implantable pacemaker, rate-responsive			E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
17949	2025	All other reportable incidents	47265	Dual-chamber implantable pacemaker, rate-responsive			E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
17950	2025	All other reportable incidents	47265	Dual-chamber implantable pacemaker, rate-responsive			E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
17951	2025	All other reportable incidents	47265	Dual-chamber implantable pacemaker, rate-responsive			E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	17/02/2026	17/02/2026
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	17/02/2026	17/02/2026
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	17/02/2026	17/02/2026
17952	2025	All other reportable incidents	47267	Single-chamber implantable pacemaker, rate-responsive			E2403	No Clinical Signs, Symptoms or Conditions	A04	Material Integrity Problem		
							F26	No Health Consequences or Impact	A04	Material Integrity Problem		
17953	2025		37265	Dual-chamber implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A0722	Impedance Problem		
									A072201	High Impedance		
							F1901	Additional Surgery	A0722	Impedance Problem		
									A072201	High Impedance		
							F2301	Additional Device Required	A0722	Impedance Problem		
									A072201	High Impedance		
17954	2025	All other reportable incidents	47265	Dual-chamber implantable pacemaker, rate-responsive			E2403	No Clinical Signs, Symptoms or Conditions	A070101	Failure to Capture		
									A070909	Oversensing		
									A0712	Pacing Problem		
									A0713	Defibrillation/Stimulation Problem		
							F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A070101	Failure to Capture		
									A070909	Oversensing		
									A0712	Pacing Problem		
									A0713	Defibrillation/Stimulation Problem		
17955	2025	All other reportable incidents			CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A071201	Failure to Convert Rhythm		
									A072201	High Impedance		
							F08	Hospitalization or Prolonged Hospitalization	A071201	Failure to Convert Rhythm		
									A072201	High Impedance		
17956	2026	All other reportable incidents	47265	Dual-chamber implantable pacemaker, rate-responsive			E2403	No Clinical Signs, Symptoms or Conditions	A070909	Oversensing		
									A0712	Pacing Problem		
							F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A070909	Oversensing		
									A0712	Pacing Problem		
17957	2026	All other reportable incidents	47265	Dual-chamber implantable pacemaker, rate-responsive			E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
17958	2026	All other reportable incidents	47267	Single-chamber implantable pacemaker, rate-responsive			E2403	No Clinical Signs, Symptoms or Conditions	A070909	Oversensing		
									A0712	Pacing Problem		
									A072201	High Impedance		
									A090801	Signal Artifact/Noise		
									A23	Use of Device Problem		
							F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A070909	Oversensing		
									A0712	Pacing Problem		
									A072201	High Impedance		
									A090801	Signal Artifact/Noise		
									A23	Use of Device Problem		
17959	2026	All other reportable incidents	47265	Dual-chamber implantable pacemaker, rate-responsive			E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
17960	2026	All other reportable incidents	47265	Dual-chamber implantable pacemaker, rate-responsive			E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
17961	2026	All other reportable incidents	47267	Single-chamber implantable pacemaker, rate-responsive			E2403	No Clinical Signs, Symptoms or Conditions	A09	Output Problem		
							F26	No Health Consequences or Impact	A09	Output Problem		
17962	2026	All other reportable incidents	47265	Dual-chamber implantable pacemaker, rate-responsive			E2403	No Clinical Signs, Symptoms or Conditions	A0701	Capturing Problem		
							F26	No Health Consequences or Impact	A0701	Capturing Problem		
17963	2026	All other reportable incidents	47267	Single-chamber implantable pacemaker, rate-responsive			E2403	No Clinical Signs, Symptoms or Conditions	A090805	Incorrect Measurement		
							F26	No Health Consequences or Impact	A090805	Incorrect Measurement		
17964	2026	All other reportable incidents	47265	Dual-chamber implantable pacemaker, rate-responsive			E2403	No Clinical Signs, Symptoms or Conditions	A071102	Failure to Interrogate		
							F26	No Health Consequences or Impact	A071102	Failure to Interrogate		
17965	2026	All other reportable incidents	47265	Dual-chamber implantable pacemaker, rate-responsive			E2403	No Clinical Signs, Symptoms or Conditions	A071102	Failure to Interrogate		
							F26	No Health Consequences or Impact	A071102	Failure to Interrogate		
17966	2026	All other reportable incidents	47265	Dual-chamber implantable pacemaker, rate-responsive			E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
17967	2026	All other reportable incidents	47265	Dual-chamber implantable pacemaker, rate-responsive			E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
17968	2026	All other reportable incidents	47265	Dual-chamber implantable pacemaker, rate-responsive			E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	01/03/2026	01/03/2026
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	01/03/2026	01/03/2026
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	01/03/2026	01/03/2026
17969	2026	All other reportable incidents	47265	Dual-chamber implantable pacemaker, rate-responsive			E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
17970	2026	All other reportable incidents	47265	Dual-chamber implantable pacemaker, rate-responsive			E2403	No Clinical Signs, Symptoms or Conditions	A071102	Failure to Interrogate		
							F26	No Health Consequences or Impact	A071102	Failure to Interrogate		
17971	2026	All other reportable incidents	47265	Dual-chamber implantable pacemaker, rate-responsive			E2403	No Clinical Signs, Symptoms or Conditions	A090811	Unable to Obtain Readings		
							F23	Unexpected Medical Intervention	A090811	Unable to Obtain Readings		
17972	2026	All other reportable incidents	47265	Dual-chamber implantable pacemaker, rate-responsive			E2403	No Clinical Signs, Symptoms or Conditions	A090801	Signal Artifact/Noise		
							F28	Appropriate Health Impact Term/Code not Available	A090801	Signal Artifact/Noise		
17973	2026	All other reportable incidents	47267	Single-chamber implantable pacemaker, rate-responsive			E2403	No Clinical Signs, Symptoms or Conditions	A09	Output Problem		
							F26	No Health Consequences or Impact	A09	Output Problem		
17974	2026	All other reportable incidents	47265	Dual-chamber implantable pacemaker, rate-responsive			E2403	No Clinical Signs, Symptoms or Conditions	A13	Communication or Transmission Problem		
							F26	No Health Consequences or Impact	A13	Communication or Transmission Problem		
17975	2026	All other reportable incidents	47267	Single-chamber implantable pacemaker, rate-responsive	CT1685	Cardiac pacemakers	E0112	Dizziness	A27	Appropriate Device Problem Term/Code not Available		
							E0717	Dyspnea	A27	Appropriate Device Problem Term/Code not Available		
							F11	Non-Serious Injury/ Illness/ Impairment	A27	Appropriate Device Problem Term/Code not Available		
					CT257	Implantable cardiac pacemakers	E0112	Dizziness	A27	Appropriate Device Problem Term/Code not Available		
							E0717	Dyspnea	A27	Appropriate Device Problem Term/Code not Available		

							F11	Non-Serious Injury/ Illness/ Impairment	A27	Appropriate Device Problem Term/Code not Available		
					CT2866	Implantable cardiac electrical stimulators	E0112	Dizziness	A27	Appropriate Device Problem Term/Code not Available		
							E0717	Dyspnea	A27	Appropriate Device Problem Term/Code not Available		
							F11	Non-Serious Injury/ Illness/ Impairment	A27	Appropriate Device Problem Term/Code not Available		
17976	2026	All other reportable incidents	47265	Dual-chamber implantable pacemaker, rate-responsive			E2403	No Clinical Signs, Symptoms or Conditions	A071102	Failure to Interrogate		
							F26	No Health Consequences or Impact	A071102	Failure to Interrogate		
17977	2026	All other reportable incidents	47265	Dual-chamber implantable pacemaker, rate-responsive			E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
17978	2026	All other reportable incidents	47265	Dual-chamber implantable pacemaker, rate-responsive			E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
17979	2026	All other reportable incidents	47263	Cardiac resynchronization therapy implantable pacemaker	CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	09/01/2026	09/01/2026
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	09/01/2026	09/01/2026
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	09/01/2026	09/01/2026
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	09/01/2026	09/01/2026
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	09/01/2026	09/01/2026
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	09/01/2026	09/01/2026
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	09/01/2026	09/01/2026
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	09/01/2026	09/01/2026
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	09/01/2026	09/01/2026
17980	2026	All other reportable incidents	47265	Dual-chamber implantable pacemaker, rate-responsive			E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem		
							F26	No Health Consequences or Impact	A0705	Battery Problem		
17981	2026	All other reportable incidents	47265	Dual-chamber implantable pacemaker, rate-responsive			E2403	No Clinical Signs, Symptoms or Conditions	A090811	Unable to Obtain Readings		
							F26	No Health Consequences or Impact	A090811	Unable to Obtain Readings		
17982	2026	All other reportable incidents	47265	Dual-chamber implantable pacemaker, rate-responsive			E2403	No Clinical Signs, Symptoms or Conditions	A05	Mechanical Problem	11/07/2022	11/07/2022
							F26	No Health Consequences or Impact	A05	Mechanical Problem	11/07/2022	11/07/2022
17983	2026	All other reportable incidents	47265	Dual-chamber implantable pacemaker, rate-responsive			E2403	No Clinical Signs, Symptoms or Conditions	A1107	Data Problem		
							F26	No Health Consequences or Impact	A1107	Data Problem		
17984	2026	All other reportable incidents	47265	Dual-chamber implantable pacemaker, rate-responsive			E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	25/02/2026	25/02/2026
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	25/02/2026	25/02/2026
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	25/02/2026	25/02/2026
17985	2026	All other reportable incidents	47265	Dual-chamber implantable pacemaker, rate-responsive			E2403	No Clinical Signs, Symptoms or Conditions	A071102	Failure to Interrogate	30/12/2025	30/12/2025
									A071204	No Pacing	30/12/2025	30/12/2025
							F1901	Additional Surgery	A071102	Failure to Interrogate	30/12/2025	30/12/2025
									A071204	No Pacing	30/12/2025	30/12/2025
17986	2026	Unanticipated serious deterioration in state of health	47267	Single-chamber implantable pacemaker, rate-responsive			E2401	Insufficient Information	A24	No Apparent Device Problem		
							F02	Death	A24	No Apparent Device Problem		
17987	2026	All other reportable incidents	47265	Dual-chamber implantable pacemaker, rate-responsive			E2403	No Clinical Signs, Symptoms or Conditions	A1605	Reset Problem		
							F26	No Health Consequences or Impact	A1605	Reset Problem		
17988	2026	All other reportable incidents	47265	Dual-chamber implantable pacemaker, rate-responsive			E2403	No Clinical Signs, Symptoms or Conditions	A071102	Failure to Interrogate		
							F26	No Health Consequences or Impact	A071102	Failure to Interrogate		
17989	2026	All other reportable incidents	47265	Dual-chamber implantable pacemaker, rate-responsive			E2403	No Clinical Signs, Symptoms or Conditions	A071102	Failure to Interrogate		
							F26	No Health Consequences or Impact	A071102	Failure to Interrogate		
17990	2026	All other reportable incidents	47265	Dual-chamber implantable pacemaker, rate-responsive			E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	04/05/2026	04/05/2026
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	04/05/2026	04/05/2026
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	04/05/2026	04/05/2026
17991	2026	All other reportable incidents	47265	Dual-chamber implantable pacemaker, rate-responsive			E2403	No Clinical Signs, Symptoms or Conditions	A0908	Incorrect, Inadequate or Imprecise Result or Readings	05/12/2025	05/12/2025
							F08	Hospitalization or Prolonged Hospitalization	A0908	Incorrect, Inadequate or Imprecise Result or Readings	05/12/2025	05/12/2025
							F1905	Device Revision or Replacement	A0908	Incorrect, Inadequate or Imprecise Result or Readings	05/12/2025	05/12/2025
17992	2026	All other reportable incidents	47265	Dual-chamber implantable pacemaker, rate-responsive			E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem		
									A0711	Interrogation Problem		
									A0908	Incorrect, Inadequate or Imprecise Result or Readings		
									A1107	Data Problem		
							F26	No Health Consequences or Impact	A0705	Battery Problem		
									A0711	Interrogation Problem		
									A0908	Incorrect, Inadequate or Imprecise Result or Readings		
									A1107	Data Problem		
17993	2026	All other reportable incidents	47265	Dual-chamber implantable pacemaker, rate-responsive			E2403	No Clinical Signs, Symptoms or Conditions	A0908	Incorrect, Inadequate or Imprecise Result or Readings	31/05/2026	31/05/2026
							F08	Hospitalization or Prolonged Hospitalization	A0908	Incorrect, Inadequate or Imprecise Result or Readings	31/05/2026	31/05/2026
							F1905	Device Revision or Replacement	A0908	Incorrect, Inadequate or Imprecise Result or Readings	31/05/2026	31/05/2026
17994	2026	All other reportable incidents	47267	Single-chamber implantable pacemaker, rate-responsive			E2403	No Clinical Signs, Symptoms or Conditions	A0701	Capturing Problem	03/02/2026	03/02/2026
									A070504	Premature Discharge of Battery	03/02/2026	03/02/2026
							F08	Hospitalization or Prolonged Hospitalization	A0701	Capturing Problem	03/02/2026	03/02/2026
									A070504	Premature Discharge of Battery	03/02/2026	03/02/2026
							F12	Serious Injury/ Illness/ Impairment	A0701	Capturing Problem	03/02/2026	03/02/2026
									A070504	Premature Discharge of Battery	03/02/2026	03/02/2026
							F1905	Device Revision or Replacement	A0701	Capturing Problem	03/02/2026	03/02/2026
									A070504	Premature Discharge of Battery	03/02/2026	03/02/2026
							F26	No Health Consequences or Impact	A0701	Capturing Problem	03/02/2026	03/02/2026
									A070504	Premature Discharge of Battery	03/02/2026	03/02/2026
17995	2026	All other reportable incidents	47265	Dual-chamber implantable pacemaker, rate-responsive			E2403	No Clinical Signs, Symptoms or Conditions	A071102	Failure to Interrogate		
							F26	No Health Consequences or Impact	A071102	Failure to Interrogate		
17996	2026	All other reportable incidents	47265	Dual-chamber implantable pacemaker, rate-responsive			E2403	No Clinical Signs, Symptoms or Conditions	A160601	Premature Elective Replacement Indicator		
							F26	No Health Consequences or Impact	A160601	Premature Elective Replacement Indicator		
17997	2026	All other reportable incidents	47265	Dual-chamber implantable pacemaker, rate-responsive			E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
17998	2026	All other reportable incidents	47265	Dual-chamber implantable pacemaker, rate-responsive			E2403	No Clinical Signs, Symptoms or Conditions	A0908	Incorrect, Inadequate or Imprecise Result or Readings		
							F26	No Health Consequences or Impact	A0908	Incorrect, Inadequate or Imprecise Result or Readings		

17999	2026	All other reportable incidents	47265	Dual-chamber implantable pacemaker, rate-responsive			E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
18000	2026	Unanticipated serious deterioration in state of health	60789	Intracardiac pacemaker			E060109	Tachycardia	A27	Appropriate Device Problem Term/Code not Available	13/01/2026	13/01/2026
							E0602	Cardiac Arrest	A27	Appropriate Device Problem Term/Code not Available	13/01/2026	13/01/2026
							F08	Hospitalization or Prolonged Hospitalization	A27	Appropriate Device Problem Term/Code not Available	13/01/2026	13/01/2026
							F1203	Life Threatening Illness or Injury	A27	Appropriate Device Problem Term/Code not Available	13/01/2026	13/01/2026
							F1905	Device Revision or Replacement	A27	Appropriate Device Problem Term/Code not Available	13/01/2026	13/01/2026
							F2307	Reprogramming of Device	A27	Appropriate Device Problem Term/Code not Available	13/01/2026	13/01/2026
18001	2026	All other reportable incidents	47267	Single-chamber implantable pacemaker, rate-responsive			E2403	No Clinical Signs, Symptoms or Conditions	A13	Communication or Transmission Problem		
							F26	No Health Consequences or Impact	A13	Communication or Transmission Problem		
18002	2026	All other reportable incidents	47265	Dual-chamber implantable pacemaker, rate-responsive			E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	23/02/2026	23/02/2026
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	23/02/2026	23/02/2026
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	23/02/2026	23/02/2026
18003	2026		47265	Dual-chamber implantable pacemaker, rate-responsive	CT2750	Implantable cardiac electrical stimulators and associated devices	E012207	Twitching	A070504	Premature Discharge of Battery		
							F01	Change in Therapeutic Response	A070504	Premature Discharge of Battery		
18004	2026	All other reportable incidents	47265	Dual-chamber implantable pacemaker, rate-responsive			E2403	No Clinical Signs, Symptoms or Conditions	A0711	Interrogation Problem	26/01/2026	26/01/2026
							F08	Hospitalization or Prolonged Hospitalization	A0711	Interrogation Problem	26/01/2026	26/01/2026
							F1905	Device Revision or Replacement	A0711	Interrogation Problem	26/01/2026	26/01/2026
18005	2026	All other reportable incidents	47265	Dual-chamber implantable pacemaker, rate-responsive			E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
18006	2026	All other reportable incidents	47265	Dual-chamber implantable pacemaker, rate-responsive			E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	27/02/2026	27/02/2026
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	27/02/2026	27/02/2026
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	27/02/2026	27/02/2026
18007	2026	All other reportable incidents	60789	Intracardiac pacemaker			E2403	No Clinical Signs, Symptoms or Conditions	A09	Output Problem		
									A13	Communication or Transmission Problem		
							F26	No Health Consequences or Impact	A09	Output Problem		
									A13	Communication or Transmission Problem		
18008	2026	All other reportable incidents	47265	Dual-chamber implantable pacemaker, rate-responsive			E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	05/03/2026	05/03/2026
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	05/03/2026	05/03/2026
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	05/03/2026	05/03/2026
18009	2026	All other reportable incidents	47265	Dual-chamber implantable pacemaker, rate-responsive			E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
18010	2026	All other reportable incidents	47265	Dual-chamber implantable pacemaker, rate-responsive			E2403	No Clinical Signs, Symptoms or Conditions	A0908	Incorrect, Inadequate or Imprecise Result or Readings	09/02/2026	09/02/2026
							F08	Hospitalization or Prolonged Hospitalization	A0908	Incorrect, Inadequate or Imprecise Result or Readings	09/02/2026	09/02/2026
							F1905	Device Revision or Replacement	A0908	Incorrect, Inadequate or Imprecise Result or Readings	09/02/2026	09/02/2026
18011	2026	All other reportable incidents	47265	Dual-chamber implantable pacemaker, rate-responsive			E2403	No Clinical Signs, Symptoms or Conditions	A0908	Incorrect, Inadequate or Imprecise Result or Readings	07/01/2026	07/01/2026
							F08	Hospitalization or Prolonged Hospitalization	A0908	Incorrect, Inadequate or Imprecise Result or Readings	07/01/2026	07/01/2026
							F1905	Device Revision or Replacement	A0908	Incorrect, Inadequate or Imprecise Result or Readings	07/01/2026	07/01/2026
18012	2026	All other reportable incidents	47265	Dual-chamber implantable pacemaker, rate-responsive			E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
18013	2026	All other reportable incidents	63188	Left-ventricular wireless pacing system sensor/transmitter			E0623	Unspecified Heart Problem	A25	No Apparent Adverse Event		
							F26	No Health Consequences or Impact	A25	No Apparent Adverse Event		
18014	2026	All other reportable incidents	47265	Dual-chamber implantable pacemaker, rate-responsive			E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	17/02/2026	17/02/2026
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	17/02/2026	17/02/2026
							F12	Serious Injury/ Illness/ Impairment	A070504	Premature Discharge of Battery	17/02/2026	17/02/2026
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	17/02/2026	17/02/2026
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery	17/02/2026	17/02/2026
18015	2026	All other reportable incidents	47265	Dual-chamber implantable pacemaker, rate-responsive			E2403	No Clinical Signs, Symptoms or Conditions	A12	Connection Problem	12/01/2026	12/01/2026
									A150206	Difficult to Insert	12/01/2026	12/01/2026
							F1908	Prolonged Surgery	A12	Connection Problem	12/01/2026	12/01/2026
									A150206	Difficult to Insert	12/01/2026	12/01/2026
18016	2026	All other reportable incidents	47265	Dual-chamber implantable pacemaker, rate-responsive	CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
									A27	Appropriate Device Problem Term/Code not Available		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
									A27	Appropriate Device Problem Term/Code not Available		
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
									A27	Appropriate Device Problem Term/Code not Available		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
									A27	Appropriate Device Problem Term/Code not Available		
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
									A27	Appropriate Device Problem Term/Code not Available		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
									A27	Appropriate Device Problem Term/Code not Available		
18017	2026	All other reportable incidents	47265	Dual-chamber implantable pacemaker, rate-responsive	CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
									A27	Appropriate Device Problem Term/Code not Available		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
									A27	Appropriate Device Problem Term/Code not Available		
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
									A27	Appropriate Device Problem Term/Code not Available		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
									A27	Appropriate Device Problem Term/Code not Available		

					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
									A27	Appropriate Device Problem Term/Code not Available		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
									A27	Appropriate Device Problem Term/Code not Available		
18018	2026	All other reportable incidents	47265	Dual-chamber implantable pacemaker, rate-responsive			E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
18019	2026	All other reportable incidents	47265	Dual-chamber implantable pacemaker, rate-responsive			E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
18020	2026	All other reportable incidents	47265	Dual-chamber implantable pacemaker, rate-responsive			E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
18021	2026	All other reportable incidents	47265	Dual-chamber implantable pacemaker, rate-responsive			E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
18022	2026		47265	Dual-chamber implantable pacemaker, rate-responsive			E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	10/01/2026	10/01/2026
							F1901	Additional Surgery	A24	No Apparent Device Problem	10/01/2026	10/01/2026
18023	2026	Unanticipated serious deterioration in state of health	47265	Dual-chamber implantable pacemaker, rate-responsive			E060101	Asystole	A0712	Pacing Problem	20/01/2026	20/01/2026
							E2403	No Clinical Signs, Symptoms or Conditions	A0712	Pacing Problem	20/01/2026	20/01/2026
							F12	Serious Injury/ Illness/ Impairment	A0712	Pacing Problem	20/01/2026	20/01/2026
							F1203	Life Threatening Illness or Injury	A0712	Pacing Problem	20/01/2026	20/01/2026
							F1905	Device Revision or Replacement	A0712	Pacing Problem	20/01/2026	20/01/2026
							F23	Unexpected Medical Intervention	A0712	Pacing Problem	20/01/2026	20/01/2026
18024	2026		47265	Dual-chamber implantable pacemaker, rate-responsive	CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem		
									A1108	Date/Time-Related Software Problem		
							F23	Unexpected Medical Intervention	A0705	Battery Problem		
									A1108	Date/Time-Related Software Problem		
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem		
									A1108	Date/Time-Related Software Problem		
							F23	Unexpected Medical Intervention	A0705	Battery Problem		
									A1108	Date/Time-Related Software Problem		
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem		
									A1108	Date/Time-Related Software Problem		
							F23	Unexpected Medical Intervention	A0705	Battery Problem		
									A1108	Date/Time-Related Software Problem		
18025	2026	Unanticipated serious deterioration in state of health	47265	Dual-chamber implantable pacemaker, rate-responsive			E060101	Asystole	A0712	Pacing Problem	21/01/2026	21/01/2026
							F1203	Life Threatening Illness or Injury	A0712	Pacing Problem	21/01/2026	21/01/2026
18026	2026	All other reportable incidents	47265	Dual-chamber implantable pacemaker, rate-responsive			E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
18027	2026	All other reportable incidents	47265	Dual-chamber implantable pacemaker, rate-responsive			E2403	No Clinical Signs, Symptoms or Conditions	A070102	High Capture Threshold		
									A070909	Oversensing		
									A072202	Low Impedance		
									A090801	Signal Artifact/Noise		
									A1301	Failure to Read Input Signal		
							F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A070102	High Capture Threshold		
									A070909	Oversensing		
									A072202	Low Impedance		
									A090801	Signal Artifact/Noise		
									A1301	Failure to Read Input Signal		
18028	2026	All other reportable incidents	47265	Dual-chamber implantable pacemaker, rate-responsive			E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
18029	2026	All other reportable incidents	47265	Dual-chamber implantable pacemaker, rate-responsive			E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
18030	2026	All other reportable incidents	47265	Dual-chamber implantable pacemaker, rate-responsive	CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
					CT2750	Implantable cardiac electrical stimulators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
18031	2026		17882	Physiologic-monitoring defibrillation system	CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A09	Output Problem		
									A191002	Loss of Power		
							F26	No Health Consequences or Impact	A09	Output Problem		
									A191002	Loss of Power		
18032	2026	All other reportable incidents	60789	Intracardiac pacemaker			E2403	No Clinical Signs, Symptoms or Conditions	A070910	Undersensing		
							F12	Serious Injury/ Illness/ Impairment	A070910	Undersensing		
							F2307	Reprogramming of Device	A070910	Undersensing		
18033	2026	All other reportable incidents	47265	Dual-chamber implantable pacemaker, rate-responsive			E2403	No Clinical Signs, Symptoms or Conditions	A0908	Incorrect, Inadequate or Imprecise Result or Readings		
							F26	No Health Consequences or Impact	A0908	Incorrect, Inadequate or Imprecise Result or Readings		
18034	2026	All other reportable incidents	47267	Single-chamber implantable pacemaker, rate-responsive			E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
18035	2026	All other reportable incidents	47265	Dual-chamber implantable pacemaker, rate-responsive			E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
18036	2026	All other reportable incidents	47265	Dual-chamber implantable pacemaker, rate-responsive			E2103	Device Overstimulation of Tissue	A0908	Incorrect, Inadequate or Imprecise Result or Readings		
							F26	No Health Consequences or Impact	A0908	Incorrect, Inadequate or Imprecise Result or Readings		

18037	2026	All other reportable incidents	47267	Single-chamber implantable pacemaker, rate-responsive			E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	05/02/2026	05/02/2026
									A090404	Output Above Specifications	05/02/2026	05/02/2026
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	05/02/2026	05/02/2026
									A090404	Output Above Specifications	05/02/2026	05/02/2026
							F12	Serious Injury/ Illness/ Impairment	A070504	Premature Discharge of Battery	05/02/2026	05/02/2026
									A090404	Output Above Specifications	05/02/2026	05/02/2026
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	05/02/2026	05/02/2026
									A090404	Output Above Specifications	05/02/2026	05/02/2026
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery	05/02/2026	05/02/2026
									A090404	Output Above Specifications	05/02/2026	05/02/2026
18038	2026	All other reportable incidents	47265	Dual-chamber implantable pacemaker, rate-responsive			E2403	No Clinical Signs, Symptoms or Conditions	A0908	Incorrect, Inadequate or Imprecise Result or Readings	16/01/2026	16/01/2026
							F08	Hospitalization or Prolonged Hospitalization	A0908	Incorrect, Inadequate or Imprecise Result or Readings	16/01/2026	16/01/2026
							F1905	Device Revision or Replacement	A0908	Incorrect, Inadequate or Imprecise Result or Readings	16/01/2026	16/01/2026
18039	2026	All other reportable incidents	47265	Dual-chamber implantable pacemaker, rate-responsive			E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem		
									A0711	Interrogation Problem		
									A0908	Incorrect, Inadequate or Imprecise Result or Readings		
									A1107	Data Problem		
							F26	No Health Consequences or Impact	A0705	Battery Problem		
									A0711	Interrogation Problem		
									A0908	Incorrect, Inadequate or Imprecise Result or Readings		
									A1107	Data Problem		
18040	2026	All other reportable incidents	47265	Dual-chamber implantable pacemaker, rate-responsive			E060106	Heart Block	A070102	High Capture Threshold		
									A070909	Oversensing		
									A072202	Low Impedance		
							F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A070102	High Capture Threshold		
									A070909	Oversensing		
									A072202	Low Impedance		
18041	2026	All other reportable incidents	47265	Dual-chamber implantable pacemaker, rate-responsive			E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
18042	2026	All other reportable incidents	47265	Dual-chamber implantable pacemaker, rate-responsive			E2403	No Clinical Signs, Symptoms or Conditions	A072201	High Impedance		
									A090801	Signal Artifact/Noise		
							F2203	Imaging Required	A072201	High Impedance		
									A090801	Signal Artifact/Noise		
							F26	No Health Consequences or Impact	A072201	High Impedance		
									A090801	Signal Artifact/Noise		
18043	2026	All other reportable incidents	47265	Dual-chamber implantable pacemaker, rate-responsive			E0109	Convulsion/Seizure	A26	Insufficient Device Problem Information	18/11/2025	18/11/2025
							F08	Hospitalization or Prolonged Hospitalization	A26	Insufficient Device Problem Information	18/11/2025	18/11/2025
							F1905	Device Revision or Replacement	A26	Insufficient Device Problem Information	18/11/2025	18/11/2025
18044	2026	All other reportable incidents	47263	Cardiac resynchronization therapy implantable pacemaker	CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
							F1905	Device Revision or Replacement	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
							F2301	Additional Device Required	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
							F1905	Device Revision or Replacement	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
							F2301	Additional Device Required	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
							F1905	Device Revision or Replacement	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
							F2301	Additional Device Required	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
18045	2026	All other reportable incidents	47265	Dual-chamber implantable pacemaker, rate-responsive			E2403	No Clinical Signs, Symptoms or Conditions	A26	Insufficient Device Problem Information		
							F26	No Health Consequences or Impact	A26	Insufficient Device Problem Information		
18046	2026	All other reportable incidents	47265	Dual-chamber implantable pacemaker, rate-responsive			E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
18047	2026	All other reportable incidents	47265	Dual-chamber implantable pacemaker, rate-responsive			E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	06/01/2026	06/01/2026
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	06/01/2026	06/01/2026
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	06/01/2026	06/01/2026
18048	2026	All other reportable incidents	47265	Dual-chamber implantable pacemaker, rate-responsive			E0119	Loss of Consciousness	A070909	Oversensing	13/01/2026	13/01/2026
									A0908	Incorrect, Inadequate or Imprecise Result or Readings	13/01/2026	13/01/2026
							F08	Hospitalization or Prolonged Hospitalization	A070909	Oversensing	13/01/2026	13/01/2026
									A0908	Incorrect, Inadequate or Imprecise Result or Readings	13/01/2026	13/01/2026
							F1905	Device Revision or Replacement	A070909	Oversensing	13/01/2026	13/01/2026
									A0908	Incorrect, Inadequate or Imprecise Result or Readings	13/01/2026	13/01/2026
18049	2026	All other reportable incidents	47263	Cardiac resynchronization therapy implantable pacemaker	CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A01	Patient Device Interaction Problem		
									A0705	Battery Problem		
									A070504	Premature Discharge of Battery		

									A0708	Power Problem		
							F26	No Health Consequences or Impact	A01	Patient Device Interaction Problem		
									A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
									A0708	Power Problem		
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A01	Patient Device Interaction Problem		
									A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
									A0708	Power Problem		
							F26	No Health Consequences or Impact	A01	Patient Device Interaction Problem		
									A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
									A0708	Power Problem		
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A01	Patient Device Interaction Problem		
									A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
									A0708	Power Problem		
							F26	No Health Consequences or Impact	A01	Patient Device Interaction Problem		
									A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
									A0708	Power Problem		
18050	2026	Death	47265	Dual-chamber implantable pacemaker, rate-responsive			E060101	Asystole	A070504	Premature Discharge of Battery	26/01/2026	26/01/2026
							E0602	Cardiac Arrest	A070504	Premature Discharge of Battery	26/01/2026	26/01/2026
							F02	Death	A070504	Premature Discharge of Battery	26/01/2026	26/01/2026
							F2306	Resuscitation	A070504	Premature Discharge of Battery	26/01/2026	26/01/2026
18051	2026		47265	Dual-chamber implantable pacemaker, rate-responsive	CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A16	Protective Measures Problem		
									A160503	Inappropriate or Unexpected Reset		
							F26	No Health Consequences or Impact	A16	Protective Measures Problem		
									A160503	Inappropriate or Unexpected Reset		
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A16	Protective Measures Problem		
									A160503	Inappropriate or Unexpected Reset		
							F26	No Health Consequences or Impact	A16	Protective Measures Problem		
									A160503	Inappropriate or Unexpected Reset		
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A16	Protective Measures Problem		
									A160503	Inappropriate or Unexpected Reset		
							F26	No Health Consequences or Impact	A16	Protective Measures Problem		
									A160503	Inappropriate or Unexpected Reset		
18052	2026	All other reportable incidents	47265	Dual-chamber implantable pacemaker, rate-responsive			E2403	No Clinical Signs, Symptoms or Conditions	A0908	Incorrect, Inadequate or Imprecise Result or Readings	20/01/2026	20/01/2026
							F1905	Device Revision or Replacement	A0908	Incorrect, Inadequate or Imprecise Result or Readings	20/01/2026	20/01/2026
18053	2026	All other reportable incidents	47265	Dual-chamber implantable pacemaker, rate-responsive			E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
18054	2026	All other reportable incidents	47265	Dual-chamber implantable pacemaker, rate-responsive			E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
18055	2026	All other reportable incidents	47265	Dual-chamber implantable pacemaker, rate-responsive			E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem	21/04/2026	21/04/2026
									A0711	Interrogation Problem	21/04/2026	21/04/2026
									A0908	Incorrect, Inadequate or Imprecise Result or Readings	21/04/2026	21/04/2026
									A1107	Data Problem	21/04/2026	21/04/2026
							F08	Hospitalization or Prolonged Hospitalization	A0705	Battery Problem	21/04/2026	21/04/2026
									A0711	Interrogation Problem	21/04/2026	21/04/2026
									A0908	Incorrect, Inadequate or Imprecise Result or Readings	21/04/2026	21/04/2026
									A1107	Data Problem	21/04/2026	21/04/2026
							F1905	Device Revision or Replacement	A0705	Battery Problem	21/04/2026	21/04/2026
									A0711	Interrogation Problem	21/04/2026	21/04/2026
									A0908	Incorrect, Inadequate or Imprecise Result or Readings	21/04/2026	21/04/2026
									A1107	Data Problem	21/04/2026	21/04/2026
18056	2026	All other reportable incidents	47265	Dual-chamber implantable pacemaker, rate-responsive			E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	19/02/2026	19/02/2026
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	19/02/2026	19/02/2026
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	19/02/2026	19/02/2026
18057	2026	All other reportable incidents	47265	Dual-chamber implantable pacemaker, rate-responsive			E0601	Arrhythmia	A0711	Interrogation Problem		
									A0712	Pacing Problem		
							F23	Unexpected Medical Intervention	A0711	Interrogation Problem		
									A0712	Pacing Problem		
18058	2026	All other reportable incidents	47265	Dual-chamber implantable pacemaker, rate-responsive			E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem		
							F26	No Health Consequences or Impact	A0705	Battery Problem		
18059	2026	All other reportable incidents	60789	Intracardiac pacemaker			E2403	No Clinical Signs, Symptoms or Conditions	A070102	High Capture Threshold	28/01/2026	28/01/2026
							F26	No Health Consequences or Impact	A070102	High Capture Threshold	28/01/2026	28/01/2026
18060	2026	All other reportable incidents	47265	Dual-chamber implantable pacemaker, rate-responsive			E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	13/02/2026	13/02/2026
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	13/02/2026	13/02/2026
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	13/02/2026	13/02/2026
18061	2026	All other reportable incidents	47265	Dual-chamber implantable pacemaker, rate-responsive			E2403	No Clinical Signs, Symptoms or Conditions	A0908	Incorrect, Inadequate or Imprecise Result or Readings	22/01/2026	22/01/2026
							F08	Hospitalization or Prolonged Hospitalization	A0908	Incorrect, Inadequate or Imprecise Result or Readings	22/01/2026	22/01/2026
							F1905	Device Revision or Replacement	A0908	Incorrect, Inadequate or Imprecise Result or Readings	22/01/2026	22/01/2026
18062	2026	All other reportable incidents	47265	Dual-chamber implantable pacemaker, rate-responsive	CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	03/02/2026	03/02/2026
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	03/02/2026	03/02/2026
							F1903	Device Explantation	A070504	Premature Discharge of Battery	03/02/2026	03/02/2026
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery	03/02/2026	03/02/2026
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	03/02/2026	03/02/2026
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	03/02/2026	03/02/2026
							F1903	Device Explantation	A070504	Premature Discharge of Battery	03/02/2026	03/02/2026
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery	03/02/2026	03/02/2026
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	03/02/2026	03/02/2026
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	03/02/2026	03/02/2026

							F1903	Device Explantation	A070504	Premature Discharge of Battery	03/02/2026	03/02/2026
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery	03/02/2026	03/02/2026
18063	2026	All other reportable incidents	47267	Single-chamber implantable pacemaker, rate-responsive			E2403	No Clinical Signs, Symptoms or Conditions	A070101	Failure to Capture	30/01/2026	30/01/2026
									A070103	Intermittent Capture	30/01/2026	30/01/2026
									A070901	Decreased Sensitivity	30/01/2026	30/01/2026
							F26	No Health Consequences or Impact	A070101	Failure to Capture	30/01/2026	30/01/2026
									A070103	Intermittent Capture	30/01/2026	30/01/2026
									A070901	Decreased Sensitivity	30/01/2026	30/01/2026
18064	2026	All other reportable incidents	47265	Dual-chamber implantable pacemaker, rate-responsive			E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
									A072202	Low Impedance		
									A090801	Signal Artifact/Noise		
							F26	No Health Consequences or Impact	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
									A072202	Low Impedance		
									A090801	Signal Artifact/Noise		
18065	2026	All other reportable incidents	47265	Dual-chamber implantable pacemaker, rate-responsive			E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
18066	2026	All other reportable incidents	47265	Dual-chamber implantable pacemaker, rate-responsive			E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem		
									A0711	Interrogation Problem		
									A0908	Incorrect, Inadequate or Imprecise Result or Readings		
									A1107	Data Problem		
							F26	No Health Consequences or Impact	A0705	Battery Problem		
									A0711	Interrogation Problem		
									A0908	Incorrect, Inadequate or Imprecise Result or Readings		
									A1107	Data Problem		
18067	2026	All other reportable incidents	47265	Dual-chamber implantable pacemaker, rate-responsive			E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
18068	2026	All other reportable incidents	47265	Dual-chamber implantable pacemaker, rate-responsive			E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
18069	2026	All other reportable incidents	47265	Dual-chamber implantable pacemaker, rate-responsive			E0112	Dizziness	A070101	Failure to Capture		
									A070103	Intermittent Capture		
									A071205	Pacemaker Found in Back-Up Mode		
									A0908	Incorrect, Inadequate or Imprecise Result or Readings		
							E0119	Loss of Consciousness	A070101	Failure to Capture		
									A070103	Intermittent Capture		
									A071205	Pacemaker Found in Back-Up Mode		
									A0908	Incorrect, Inadequate or Imprecise Result or Readings		
							E011903	Syncope/Fainting	A070101	Failure to Capture		
									A070103	Intermittent Capture		
									A071205	Pacemaker Found in Back-Up Mode		
									A0908	Incorrect, Inadequate or Imprecise Result or Readings		
							E2007	Fall	A070101	Failure to Capture		
									A070103	Intermittent Capture		
									A071205	Pacemaker Found in Back-Up Mode		
									A0908	Incorrect, Inadequate or Imprecise Result or Readings		
							E2403	No Clinical Signs, Symptoms or Conditions	A070101	Failure to Capture		
									A070103	Intermittent Capture		
									A071205	Pacemaker Found in Back-Up Mode		
									A0908	Incorrect, Inadequate or Imprecise Result or Readings		
							F08	Hospitalization or Prolonged Hospitalization	A070101	Failure to Capture		
									A070103	Intermittent Capture		
									A071205	Pacemaker Found in Back-Up Mode		
									A0908	Incorrect, Inadequate or Imprecise Result or Readings		
							F1001	Absence of Treatment	A070101	Failure to Capture		
									A070103	Intermittent Capture		
									A071205	Pacemaker Found in Back-Up Mode		
									A0908	Incorrect, Inadequate or Imprecise Result or Readings		
							F1901	Additional Surgery	A070101	Failure to Capture		
									A070103	Intermittent Capture		
									A071205	Pacemaker Found in Back-Up Mode		
									A0908	Incorrect, Inadequate or Imprecise Result or Readings		
							F1905	Device Revision or Replacement	A070101	Failure to Capture		
									A070103	Intermittent Capture		
									A071205	Pacemaker Found in Back-Up Mode		
									A0908	Incorrect, Inadequate or Imprecise Result or Readings		
							F2301	Additional Device Required	A070101	Failure to Capture		
									A070103	Intermittent Capture		
									A071205	Pacemaker Found in Back-Up Mode		
									A0908	Incorrect, Inadequate or Imprecise Result or Readings		
18070	2026		47263	Cardiac resynchronization therapy implantable pacemaker	CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem		
									A090811	Unable to Obtain Readings		
							F19	Surgical Intervention	A0705	Battery Problem		
									A090811	Unable to Obtain Readings		
							F26	No Health Consequences or Impact	A0705	Battery Problem		
									A090811	Unable to Obtain Readings		
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem		
									A090811	Unable to Obtain Readings		
							F19	Surgical Intervention	A0705	Battery Problem		

									A090811	Unable to Obtain Readings		
							F26	No Health Consequences or Impact	A0705	Battery Problem		
									A090811	Unable to Obtain Readings		
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem		
									A090811	Unable to Obtain Readings		
							F19	Surgical Intervention	A0705	Battery Problem		
									A090811	Unable to Obtain Readings		
							F26	No Health Consequences or Impact	A0705	Battery Problem		
									A090811	Unable to Obtain Readings		
18071	2026	All other reportable incidents	47265	Dual-chamber implantable pacemaker, rate-responsive			E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem	02/02/2026	02/02/2026
									A0711	Interrogation Problem	02/02/2026	02/02/2026
									A0908	Incorrect, Inadequate or Imprecise Result or Readings	02/02/2026	02/02/2026
							F08	Hospitalization or Prolonged Hospitalization	A0705	Battery Problem	02/02/2026	02/02/2026
									A0711	Interrogation Problem	02/02/2026	02/02/2026
									A0908	Incorrect, Inadequate or Imprecise Result or Readings	02/02/2026	02/02/2026
							F1905	Device Revision or Replacement	A0705	Battery Problem	02/02/2026	02/02/2026
									A0711	Interrogation Problem	02/02/2026	02/02/2026
									A0908	Incorrect, Inadequate or Imprecise Result or Readings	02/02/2026	02/02/2026
18072	2026	All other reportable incidents	47265	Dual-chamber implantable pacemaker, rate-responsive			E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	02/03/2026	02/03/2026
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	02/03/2026	02/03/2026
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	02/03/2026	02/03/2026
18073	2026	All other reportable incidents	47265	Dual-chamber implantable pacemaker, rate-responsive			E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	22/01/2026	22/01/2026
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	22/01/2026	22/01/2026
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	22/01/2026	22/01/2026
18074	2026	All other reportable incidents	47267	Single-chamber implantable pacemaker, rate-responsive			E2403	No Clinical Signs, Symptoms or Conditions	A05	Mechanical Problem	04/02/2026	04/02/2026
									A0712	Pacing Problem	04/02/2026	04/02/2026
							F26	No Health Consequences or Impact	A05	Mechanical Problem	04/02/2026	04/02/2026
									A0712	Pacing Problem	04/02/2026	04/02/2026
18075	2026	All other reportable incidents	47265	Dual-chamber implantable pacemaker, rate-responsive	CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
									A16	Protective Measures Problem		
									A26	Insufficient Device Problem Information		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
									A16	Protective Measures Problem		
									A26	Insufficient Device Problem Information		
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
									A16	Protective Measures Problem		
									A26	Insufficient Device Problem Information		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
									A16	Protective Measures Problem		
									A26	Insufficient Device Problem Information		
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
									A16	Protective Measures Problem		
									A26	Insufficient Device Problem Information		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
									A16	Protective Measures Problem		
									A26	Insufficient Device Problem Information		
18076	2026	All other reportable incidents	47263	Cardiac resynchronization therapy implantable pacemaker	CT1685	Cardiac pacemakers	E010701	Confusion/ Disorientation	A070909	Oversensing	21/01/2026	21/01/2026
									A0712	Pacing Problem	21/01/2026	21/01/2026
									A071205	Pacemaker Found in Back-Up Mode	21/01/2026	21/01/2026
									A090406	Therapeutic or Diagnostic Output Failure	21/01/2026	21/01/2026
									A0908	Incorrect, Inadequate or Imprecise Result or Readings	21/01/2026	21/01/2026
									A1602	Fail-Safe Problem	21/01/2026	21/01/2026
							E011903	Syncope/Fainting	A070909	Oversensing	21/01/2026	21/01/2026
									A0712	Pacing Problem	21/01/2026	21/01/2026
									A071205	Pacemaker Found in Back-Up Mode	21/01/2026	21/01/2026
									A090406	Therapeutic or Diagnostic Output Failure	21/01/2026	21/01/2026
									A0908	Incorrect, Inadequate or Imprecise Result or Readings	21/01/2026	21/01/2026
									A1602	Fail-Safe Problem	21/01/2026	21/01/2026
							E1610	Head Injury	A070909	Oversensing	21/01/2026	21/01/2026
									A0712	Pacing Problem	21/01/2026	21/01/2026
									A071205	Pacemaker Found in Back-Up Mode	21/01/2026	21/01/2026
									A090406	Therapeutic or Diagnostic Output Failure	21/01/2026	21/01/2026
									A0908	Incorrect, Inadequate or Imprecise Result or Readings	21/01/2026	21/01/2026
									A1602	Fail-Safe Problem	21/01/2026	21/01/2026
							F08	Hospitalization or Prolonged Hospitalization	A070909	Oversensing	21/01/2026	21/01/2026
									A0712	Pacing Problem	21/01/2026	21/01/2026
									A071205	Pacemaker Found in Back-Up Mode	21/01/2026	21/01/2026
									A090406	Therapeutic or Diagnostic Output Failure	21/01/2026	21/01/2026
									A0908	Incorrect, Inadequate or Imprecise Result or Readings	21/01/2026	21/01/2026
									A1602	Fail-Safe Problem	21/01/2026	21/01/2026
							F1001	Absence of Treatment	A070909	Oversensing	21/01/2026	21/01/2026
									A0712	Pacing Problem	21/01/2026	21/01/2026
									A071205	Pacemaker Found in Back-Up Mode	21/01/2026	21/01/2026
									A090406	Therapeutic or Diagnostic Output Failure	21/01/2026	21/01/2026
									A0908	Incorrect, Inadequate or Imprecise Result or Readings	21/01/2026	21/01/2026
									A1602	Fail-Safe Problem	21/01/2026	21/01/2026
							F12	Serious Injury/ Illness/ Impairment	A070909	Oversensing	21/01/2026	21/01/2026
									A0712	Pacing Problem	21/01/2026	21/01/2026
									A071205	Pacemaker Found in Back-Up Mode	21/01/2026	21/01/2026
									A090406	Therapeutic or Diagnostic Output Failure	21/01/2026	21/01/2026
									A0908	Incorrect, Inadequate or Imprecise Result or Readings	21/01/2026	21/01/2026

								A1602	Fail-Safe Problem	21/01/2026	21/01/2026
						F1905	Device Revision or Replacement	A070909	Oversensing	21/01/2026	21/01/2026
								A0712	Pacing Problem	21/01/2026	21/01/2026
								A071205	Pacemaker Found in Back-Up Mode	21/01/2026	21/01/2026
								A090406	Therapeutic or Diagnostic Output Failure	21/01/2026	21/01/2026
								A0908	Incorrect, Inadequate or Imprecise Result or Readings	21/01/2026	21/01/2026
								A1602	Fail-Safe Problem	21/01/2026	21/01/2026
					CT257	Implantable cardiac pacemakers	E010701	Confusion/ Disorientation	A070909	Oversensing	21/01/2026	21/01/2026
									A0712	Pacing Problem	21/01/2026	21/01/2026
									A071205	Pacemaker Found in Back-Up Mode	21/01/2026	21/01/2026
									A090406	Therapeutic or Diagnostic Output Failure	21/01/2026	21/01/2026
									A0908	Incorrect, Inadequate or Imprecise Result or Readings	21/01/2026	21/01/2026
									A1602	Fail-Safe Problem	21/01/2026	21/01/2026
							E011903	Syncope/Fainting	A070909	Oversensing	21/01/2026	21/01/2026
									A0712	Pacing Problem	21/01/2026	21/01/2026
									A071205	Pacemaker Found in Back-Up Mode	21/01/2026	21/01/2026
									A090406	Therapeutic or Diagnostic Output Failure	21/01/2026	21/01/2026
									A0908	Incorrect, Inadequate or Imprecise Result or Readings	21/01/2026	21/01/2026
									A1602	Fail-Safe Problem	21/01/2026	21/01/2026
							E1610	Head Injury	A070909	Oversensing	21/01/2026	21/01/2026
									A0712	Pacing Problem	21/01/2026	21/01/2026
									A071205	Pacemaker Found in Back-Up Mode	21/01/2026	21/01/2026
									A090406	Therapeutic or Diagnostic Output Failure	21/01/2026	21/01/2026
									A0908	Incorrect, Inadequate or Imprecise Result or Readings	21/01/2026	21/01/2026
									A1602	Fail-Safe Problem	21/01/2026	21/01/2026
							F08	Hospitalization or Prolonged Hospitalization	A070909	Oversensing	21/01/2026	21/01/2026
									A0712	Pacing Problem	21/01/2026	21/01/2026
									A071205	Pacemaker Found in Back-Up Mode	21/01/2026	21/01/2026
									A090406	Therapeutic or Diagnostic Output Failure	21/01/2026	21/01/2026
									A0908	Incorrect, Inadequate or Imprecise Result or Readings	21/01/2026	21/01/2026
									A1602	Fail-Safe Problem	21/01/2026	21/01/2026
							F1001	Absence of Treatment	A070909	Oversensing	21/01/2026	21/01/2026
									A0712	Pacing Problem	21/01/2026	21/01/2026
									A071205	Pacemaker Found in Back-Up Mode	21/01/2026	21/01/2026
									A090406	Therapeutic or Diagnostic Output Failure	21/01/2026	21/01/2026
									A0908	Incorrect, Inadequate or Imprecise Result or Readings	21/01/2026	21/01/2026
									A1602	Fail-Safe Problem	21/01/2026	21/01/2026
							F12	Serious Injury/ Illness/ Impairment	A070909	Oversensing	21/01/2026	21/01/2026
									A0712	Pacing Problem	21/01/2026	21/01/2026
									A071205	Pacemaker Found in Back-Up Mode	21/01/2026	21/01/2026
									A090406	Therapeutic or Diagnostic Output Failure	21/01/2026	21/01/2026
									A0908	Incorrect, Inadequate or Imprecise Result or Readings	21/01/2026	21/01/2026
									A1602	Fail-Safe Problem	21/01/2026	21/01/2026
							F1905	Device Revision or Replacement	A070909	Oversensing	21/01/2026	21/01/2026
									A0712	Pacing Problem	21/01/2026	21/01/2026
									A071205	Pacemaker Found in Back-Up Mode	21/01/2026	21/01/2026
									A090406	Therapeutic or Diagnostic Output Failure	21/01/2026	21/01/2026
									A0908	Incorrect, Inadequate or Imprecise Result or Readings	21/01/2026	21/01/2026
									A1602	Fail-Safe Problem	21/01/2026	21/01/2026
					CT2866	Implantable cardiac electrical stimulators	E010701	Confusion/ Disorientation	A070909	Oversensing	21/01/2026	21/01/2026
									A0712	Pacing Problem	21/01/2026	21/01/2026
									A071205	Pacemaker Found in Back-Up Mode	21/01/2026	21/01/2026
									A090406	Therapeutic or Diagnostic Output Failure	21/01/2026	21/01/2026
									A0908	Incorrect, Inadequate or Imprecise Result or Readings	21/01/2026	21/01/2026
									A1602	Fail-Safe Problem	21/01/2026	21/01/2026
							E011903	Syncope/Fainting	A070909	Oversensing	21/01/2026	21/01/2026
									A0712	Pacing Problem	21/01/2026	21/01/2026
									A071205	Pacemaker Found in Back-Up Mode	21/01/2026	21/01/2026
									A090406	Therapeutic or Diagnostic Output Failure	21/01/2026	21/01/2026
									A0908	Incorrect, Inadequate or Imprecise Result or Readings	21/01/2026	21/01/2026
									A1602	Fail-Safe Problem	21/01/2026	21/01/2026
							E1610	Head Injury	A070909	Oversensing	21/01/2026	21/01/2026
									A0712	Pacing Problem	21/01/2026	21/01/2026
									A071205	Pacemaker Found in Back-Up Mode	21/01/2026	21/01/2026
									A090406	Therapeutic or Diagnostic Output Failure	21/01/2026	21/01/2026
									A0908	Incorrect, Inadequate or Imprecise Result or Readings	21/01/2026	21/01/2026
									A1602	Fail-Safe Problem	21/01/2026	21/01/2026
							F08	Hospitalization or Prolonged Hospitalization	A070909	Oversensing	21/01/2026	21/01/2026
									A0712	Pacing Problem	21/01/2026	21/01/2026
									A071205	Pacemaker Found in Back-Up Mode	21/01/2026	21/01/2026
									A090406	Therapeutic or Diagnostic Output Failure	21/01/2026	21/01/2026
									A0908	Incorrect, Inadequate or Imprecise Result or Readings	21/01/2026	21/01/2026
									A1602	Fail-Safe Problem	21/01/2026	21/01/2026
							F1001	Absence of Treatment	A070909	Oversensing	21/01/2026	21/01/2026
									A0712	Pacing Problem	21/01/2026	21/01/2026
									A071205	Pacemaker Found in Back-Up Mode	21/01/2026	21/01/2026
									A090406	Therapeutic or Diagnostic Output Failure	21/01/2026	21/01/2026
									A0908	Incorrect, Inadequate or Imprecise Result or Readings	21/01/2026	21/01/2026

									A1602	Fail-Safe Problem	21/01/2026	21/01/2026
							F12	Serious Injury/ Illness/ Impairment	A070909	Oversensing	21/01/2026	21/01/2026
									A0712	Pacing Problem	21/01/2026	21/01/2026
									A071205	Pacemaker Found in Back-Up Mode	21/01/2026	21/01/2026
									A090406	Therapeutic or Diagnostic Output Failure	21/01/2026	21/01/2026
									A0908	Incorrect, Inadequate or Imprecise Result or Readings	21/01/2026	21/01/2026
									A1602	Fail-Safe Problem	21/01/2026	21/01/2026
							F1905	Device Revision or Replacement	A070909	Oversensing	21/01/2026	21/01/2026
									A0712	Pacing Problem	21/01/2026	21/01/2026
									A071205	Pacemaker Found in Back-Up Mode	21/01/2026	21/01/2026
									A090406	Therapeutic or Diagnostic Output Failure	21/01/2026	21/01/2026
									A0908	Incorrect, Inadequate or Imprecise Result or Readings	21/01/2026	21/01/2026
									A1602	Fail-Safe Problem	21/01/2026	21/01/2026
18077	2026		35852	Single-chamber implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E020201	Anxiety	A0711	Interrogation Problem		
									A2205	Device Provides Uninformative Message		
									A27	Appropriate Device Problem Term/Code not Available		
							E2403	No Clinical Signs, Symptoms or Conditions	A0711	Interrogation Problem		
									A2205	Device Provides Uninformative Message		
									A27	Appropriate Device Problem Term/Code not Available		
							F1901	Additional Surgery	A0711	Interrogation Problem		
									A2205	Device Provides Uninformative Message		
									A27	Appropriate Device Problem Term/Code not Available		
							F26	No Health Consequences or Impact	A0711	Interrogation Problem		
									A2205	Device Provides Uninformative Message		
									A27	Appropriate Device Problem Term/Code not Available		
18078	2026	All other reportable incidents	60789	Intracardiac pacemaker			E2403	No Clinical Signs, Symptoms or Conditions	A040611	Stretched	27/01/2026	27/01/2026
									A051201	Device Dislodged or Dislocated	27/01/2026	27/01/2026
									A0701	Capturing Problem	27/01/2026	27/01/2026
									A072202	Low Impedance	27/01/2026	27/01/2026
							F1901	Additional Surgery	A040611	Stretched	27/01/2026	27/01/2026
									A051201	Device Dislodged or Dislocated	27/01/2026	27/01/2026
									A0701	Capturing Problem	27/01/2026	27/01/2026
									A072202	Low Impedance	27/01/2026	27/01/2026
18079	2026	All other reportable incidents	47265	Dual-chamber implantable pacemaker, rate-responsive			E2403	No Clinical Signs, Symptoms or Conditions	A072201	High Impedance		
							F26	No Health Consequences or Impact	A072201	High Impedance		
18080	2026	All other reportable incidents	47265	Dual-chamber implantable pacemaker, rate-responsive			E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	10/04/2026	10/04/2026
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	10/04/2026	10/04/2026
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	10/04/2026	10/04/2026
18081	2026	All other reportable incidents	47265	Dual-chamber implantable pacemaker, rate-responsive			E2403	No Clinical Signs, Symptoms or Conditions	A0908	Incorrect, Inadequate or Imprecise Result or Readings		
							F26	No Health Consequences or Impact	A0908	Incorrect, Inadequate or Imprecise Result or Readings		
18082	2026	All other reportable incidents	47265	Dual-chamber implantable pacemaker, rate-responsive			E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	02/03/2026	02/03/2026
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	02/03/2026	02/03/2026
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	02/03/2026	02/03/2026
18083	2026	All other reportable incidents	47267	Single-chamber implantable pacemaker, rate-responsive			E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
18084	2026	All other reportable incidents	47265	Dual-chamber implantable pacemaker, rate-responsive			E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
18085	2026	All other reportable incidents	47265	Dual-chamber implantable pacemaker, rate-responsive			E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	30/01/2026	30/01/2026
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	30/01/2026	30/01/2026
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	30/01/2026	30/01/2026
18086	2026	All other reportable incidents	47265	Dual-chamber implantable pacemaker, rate-responsive			E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
18087	2026	All other reportable incidents	47265	Dual-chamber implantable pacemaker, rate-responsive			E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
18088	2026	All other reportable incidents	47265	Dual-chamber implantable pacemaker, rate-responsive			E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
18089	2026	All other reportable incidents	47265	Dual-chamber implantable pacemaker, rate-responsive			E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
18090	2026	All other reportable incidents	47265	Dual-chamber implantable pacemaker, rate-responsive	CT1685	Cardiac pacemakers	E2007	Fall	A0908	Incorrect, Inadequate or Imprecise Result or Readings	02/02/2026	02/02/2026
							F08	Hospitalization or Prolonged Hospitalization	A0908	Incorrect, Inadequate or Imprecise Result or Readings	02/02/2026	02/02/2026
							F1905	Device Revision or Replacement	A0908	Incorrect, Inadequate or Imprecise Result or Readings	02/02/2026	02/02/2026
					CT257	Implantable cardiac pacemakers	E2007	Fall	A0908	Incorrect, Inadequate or Imprecise Result or Readings	02/02/2026	02/02/2026
							F08	Hospitalization or Prolonged Hospitalization	A0908	Incorrect, Inadequate or Imprecise Result or Readings	02/02/2026	02/02/2026
							F1905	Device Revision or Replacement	A0908	Incorrect, Inadequate or Imprecise Result or Readings	02/02/2026	02/02/2026
					CT2866	Implantable cardiac electrical stimulators	E2007	Fall	A0908	Incorrect, Inadequate or Imprecise Result or Readings	02/02/2026	02/02/2026
							F08	Hospitalization or Prolonged Hospitalization	A0908	Incorrect, Inadequate or Imprecise Result or Readings	02/02/2026	02/02/2026
							F1905	Device Revision or Replacement	A0908	Incorrect, Inadequate or Imprecise Result or Readings	02/02/2026	02/02/2026
18091	2026	All other reportable incidents	47265	Dual-chamber implantable pacemaker, rate-responsive			E2403	No Clinical Signs, Symptoms or Conditions	A070102	High Capture Threshold		
									A070909	Oversensing		
									A070910	Undersensing		
									A072202	Low Impedance		
									A1301	Failure to Read Input Signal		
							F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A070102	High Capture Threshold		
									A070909	Oversensing		
									A070910	Undersensing		

									A072202	Low Impedance		
									A1301	Failure to Read Input Signal		
18092	2026	All other reportable incidents	47265	Dual-chamber implantable pacemaker, rate-responsive			E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
18093	2026	All other reportable incidents	47265	Dual-chamber implantable pacemaker, rate-responsive			E2403	No Clinical Signs, Symptoms or Conditions	A13	Communication or Transmission Problem		
							F26	No Health Consequences or Impact	A13	Communication or Transmission Problem		
18094	2026	All other reportable incidents	47265	Dual-chamber implantable pacemaker, rate-responsive			E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
18095	2026	All other reportable incidents	47265	Dual-chamber implantable pacemaker, rate-responsive			E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
18096	2026	All other reportable incidents	47206	Cardiac pulse generator software	CT1684	Cardiac pacemakers and associated devices	E210401	Shock from Patient Lead(s)	A070909	Oversensing		
									A070910	Undersensing		
									A071302	Inappropriate/Inadequate Shock/Stimulation		
									A1301	Failure to Read Input Signal		
							F08	Hospitalization or Prolonged Hospitalization	A070909	Oversensing		
									A070910	Undersensing		
									A071302	Inappropriate/Inadequate Shock/Stimulation		
									A1301	Failure to Read Input Signal		
							F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A070909	Oversensing		
									A070910	Undersensing		
									A071302	Inappropriate/Inadequate Shock/Stimulation		
									A1301	Failure to Read Input Signal		
							F1905	Device Revision or Replacement	A070909	Oversensing		
									A070910	Undersensing		
									A071302	Inappropriate/Inadequate Shock/Stimulation		
									A1301	Failure to Read Input Signal		
							F23	Unexpected Medical Intervention	A070909	Oversensing		
									A070910	Undersensing		
									A071302	Inappropriate/Inadequate Shock/Stimulation		
									A1301	Failure to Read Input Signal		
18097	2026	All other reportable incidents	47265	Dual-chamber implantable pacemaker, rate-responsive	CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A071102	Failure to Interrogate		
							F26	No Health Consequences or Impact	A071102	Failure to Interrogate		
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A071102	Failure to Interrogate		
							F26	No Health Consequences or Impact	A071102	Failure to Interrogate		
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A071102	Failure to Interrogate		
							F26	No Health Consequences or Impact	A071102	Failure to Interrogate		
18098	2026	Serious public health threat	17882	Physiologic-monitoring defibrillation system	CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A1603	Failure of Device to Self-Test		
							F27	Problem Identified During Non-Clinical Procedure	A1603	Failure of Device to Self-Test		
18099	2026	All other reportable incidents	47267	Single-chamber implantable pacemaker, rate-responsive			E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
18100	2026	All other reportable incidents	47265	Dual-chamber implantable pacemaker, rate-responsive			E2403	No Clinical Signs, Symptoms or Conditions	A0908	Incorrect, Inadequate or Imprecise Result or Readings	09/02/2026	09/02/2026
							F08	Hospitalization or Prolonged Hospitalization	A0908	Incorrect, Inadequate or Imprecise Result or Readings	09/02/2026	09/02/2026
							F1905	Device Revision or Replacement	A0908	Incorrect, Inadequate or Imprecise Result or Readings	09/02/2026	09/02/2026
18101	2026	All other reportable incidents	47265	Dual-chamber implantable pacemaker, rate-responsive			E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
18102	2026	All other reportable incidents	47265	Dual-chamber implantable pacemaker, rate-responsive			E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
									A1107	Data Problem		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
									A1107	Data Problem		
18103	2026	All other reportable incidents	47265	Dual-chamber implantable pacemaker, rate-responsive			E2403	No Clinical Signs, Symptoms or Conditions	A070909	Oversensing		
									A1107	Data Problem		
							F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A070909	Oversensing		
									A1107	Data Problem		
18104	2026	All other reportable incidents	47265	Dual-chamber implantable pacemaker, rate-responsive			E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
18105	2026	All other reportable incidents	47265	Dual-chamber implantable pacemaker, rate-responsive			E060104	Bradycardia	A070909	Oversensing		
									A0712	Pacing Problem		
							F1001	Absence of Treatment	A070909	Oversensing		
									A0712	Pacing Problem		
18106	2026	All other reportable incidents	47267	Single-chamber implantable pacemaker, rate-responsive			E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
18107	2026	All other reportable incidents	60789	Intracardiac pacemaker			E2403	No Clinical Signs, Symptoms or Conditions	A071102	Failure to Interrogate		
							F26	No Health Consequences or Impact	A071102	Failure to Interrogate		
18108	2026		47265	Dual-chamber implantable pacemaker, rate-responsive	CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
									A0708	Power Problem		
									A191003	Power Conditioning Problem		
							F23	Unexpected Medical Intervention	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
									A0708	Power Problem		
									A191003	Power Conditioning Problem		
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
									A0708	Power Problem		
									A191003	Power Conditioning Problem		
							F23	Unexpected Medical Intervention	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
									A0708	Power Problem		
									A191003	Power Conditioning Problem		
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
									A0708	Power Problem		
									A191003	Power Conditioning Problem		
							F23	Unexpected Medical Intervention	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		

									A0708	Power Problem		
									A191003	Power Conditioning Problem		
18109	2026	All other reportable incidents	47265	Dual-chamber implantable pacemaker, rate-responsive			E020201	Anxiety	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
18110	2026	All other reportable incidents	47265	Dual-chamber implantable pacemaker, rate-responsive			E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem		
							F26	No Health Consequences or Impact	A0705	Battery Problem		
18111	2026	All other reportable incidents	47265	Dual-chamber implantable pacemaker, rate-responsive			E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
18112	2026	All other reportable incidents	47267	Single-chamber implantable pacemaker, rate-responsive			E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	27/02/2026	27/02/2026
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	27/02/2026	27/02/2026
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	27/02/2026	27/02/2026
18113	2026	All other reportable incidents	47265	Dual-chamber implantable pacemaker, rate-responsive			E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
18114	2026	All other reportable incidents	47265	Dual-chamber implantable pacemaker, rate-responsive			E2403	No Clinical Signs, Symptoms or Conditions	A1107	Data Problem		
							F26	No Health Consequences or Impact	A1107	Data Problem		
18115	2026	All other reportable incidents	47265	Dual-chamber implantable pacemaker, rate-responsive			E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	01/04/2026	01/04/2026
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	01/04/2026	01/04/2026
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	01/04/2026	01/04/2026
18116	2026	All other reportable incidents	47265	Dual-chamber implantable pacemaker, rate-responsive			E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
18117	2026	All other reportable incidents	47265	Dual-chamber implantable pacemaker, rate-responsive			E2403	No Clinical Signs, Symptoms or Conditions	A05	Mechanical Problem		
							F26	No Health Consequences or Impact	A05	Mechanical Problem		
18118	2026	All other reportable incidents	47265	Dual-chamber implantable pacemaker, rate-responsive	CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A0201	Product Quality Problem	18/05/2026	18/05/2026
									A0705	Battery Problem	18/05/2026	18/05/2026
									A070504	Premature Discharge of Battery	18/05/2026	18/05/2026
							F08	Hospitalization or Prolonged Hospitalization	A0201	Product Quality Problem	18/05/2026	18/05/2026
									A0705	Battery Problem	18/05/2026	18/05/2026
									A070504	Premature Discharge of Battery	18/05/2026	18/05/2026
							F19	Surgical Intervention	A0201	Product Quality Problem	18/05/2026	18/05/2026
									A0705	Battery Problem	18/05/2026	18/05/2026
									A070504	Premature Discharge of Battery	18/05/2026	18/05/2026
							F1903	Device Explantation	A0201	Product Quality Problem	18/05/2026	18/05/2026
									A0705	Battery Problem	18/05/2026	18/05/2026
									A070504	Premature Discharge of Battery	18/05/2026	18/05/2026
							F26	No Health Consequences or Impact	A0201	Product Quality Problem	18/05/2026	18/05/2026
									A0705	Battery Problem	18/05/2026	18/05/2026
									A070504	Premature Discharge of Battery	18/05/2026	18/05/2026
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A0201	Product Quality Problem	18/05/2026	18/05/2026
									A0705	Battery Problem	18/05/2026	18/05/2026
									A070504	Premature Discharge of Battery	18/05/2026	18/05/2026
							F08	Hospitalization or Prolonged Hospitalization	A0201	Product Quality Problem	18/05/2026	18/05/2026
									A0705	Battery Problem	18/05/2026	18/05/2026
									A070504	Premature Discharge of Battery	18/05/2026	18/05/2026
							F19	Surgical Intervention	A0201	Product Quality Problem	18/05/2026	18/05/2026
									A0705	Battery Problem	18/05/2026	18/05/2026
									A070504	Premature Discharge of Battery	18/05/2026	18/05/2026
							F1903	Device Explantation	A0201	Product Quality Problem	18/05/2026	18/05/2026
									A0705	Battery Problem	18/05/2026	18/05/2026
									A070504	Premature Discharge of Battery	18/05/2026	18/05/2026
							F26	No Health Consequences or Impact	A0201	Product Quality Problem	18/05/2026	18/05/2026
									A0705	Battery Problem	18/05/2026	18/05/2026
									A070504	Premature Discharge of Battery	18/05/2026	18/05/2026
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A0201	Product Quality Problem	18/05/2026	18/05/2026
									A0705	Battery Problem	18/05/2026	18/05/2026
									A070504	Premature Discharge of Battery	18/05/2026	18/05/2026
							F08	Hospitalization or Prolonged Hospitalization	A0201	Product Quality Problem	18/05/2026	18/05/2026
									A0705	Battery Problem	18/05/2026	18/05/2026
									A070504	Premature Discharge of Battery	18/05/2026	18/05/2026
							F19	Surgical Intervention	A0201	Product Quality Problem	18/05/2026	18/05/2026
									A0705	Battery Problem	18/05/2026	18/05/2026
									A070504	Premature Discharge of Battery	18/05/2026	18/05/2026
							F1903	Device Explantation	A0201	Product Quality Problem	18/05/2026	18/05/2026
									A0705	Battery Problem	18/05/2026	18/05/2026
									A070504	Premature Discharge of Battery	18/05/2026	18/05/2026
							F26	No Health Consequences or Impact	A0201	Product Quality Problem	18/05/2026	18/05/2026
									A0705	Battery Problem	18/05/2026	18/05/2026
									A070504	Premature Discharge of Battery	18/05/2026	18/05/2026
18119	2026	All other reportable incidents	47265	Dual-chamber implantable pacemaker, rate-responsive			E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	17/02/2026	17/02/2026
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	17/02/2026	17/02/2026
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	17/02/2026	17/02/2026
18120	2026	All other reportable incidents	47265	Dual-chamber implantable pacemaker, rate-responsive			E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
18121	2026	All other reportable incidents	47265	Dual-chamber implantable pacemaker, rate-responsive			E2403	No Clinical Signs, Symptoms or Conditions	A1605	Reset Problem		
							F26	No Health Consequences or Impact	A1605	Reset Problem		
18122	2026	All other reportable incidents	47265	Dual-chamber implantable pacemaker, rate-responsive			E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
18123	2026		36102	Pacing/defibrillation lead electrical adaptor	CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A0202	Defective Component		
									A0203	Defective Device		
									A04	Material Integrity Problem		
									A0404	Crack		
							F26	No Health Consequences or Impact	A0202	Defective Component		
									A0203	Defective Device		
									A04	Material Integrity Problem		
									A0404	Crack		
18124	2026	All other reportable incidents	47265	Dual-chamber implantable pacemaker, rate-responsive			E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
18125	2026	All other reportable incidents	47265	Dual-chamber implantable pacemaker, rate-responsive			E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		

18126	2026	All other reportable incidents	47265	Dual-chamber implantable pacemaker, rate-responsive			E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
18127	2026	All other reportable incidents	47265	Dual-chamber implantable pacemaker, rate-responsive			E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
18128	2026	All other reportable incidents	47265	Dual-chamber implantable pacemaker, rate-responsive	CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A0908	Incorrect, Inadequate or Imprecise Result or Readings	25/02/2026	25/02/2026
							F08	Hospitalization or Prolonged Hospitalization	A0908	Incorrect, Inadequate or Imprecise Result or Readings	25/02/2026	25/02/2026
							F1905	Device Revision or Replacement	A0908	Incorrect, Inadequate or Imprecise Result or Readings	25/02/2026	25/02/2026
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A0908	Incorrect, Inadequate or Imprecise Result or Readings	25/02/2026	25/02/2026
							F08	Hospitalization or Prolonged Hospitalization	A0908	Incorrect, Inadequate or Imprecise Result or Readings	25/02/2026	25/02/2026
							F1905	Device Revision or Replacement	A0908	Incorrect, Inadequate or Imprecise Result or Readings	25/02/2026	25/02/2026
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A0908	Incorrect, Inadequate or Imprecise Result or Readings	25/02/2026	25/02/2026
							F08	Hospitalization or Prolonged Hospitalization	A0908	Incorrect, Inadequate or Imprecise Result or Readings	25/02/2026	25/02/2026
							F1905	Device Revision or Replacement	A0908	Incorrect, Inadequate or Imprecise Result or Readings	25/02/2026	25/02/2026
18129	2026	All other reportable incidents	47265	Dual-chamber implantable pacemaker, rate-responsive			E2403	No Clinical Signs, Symptoms or Conditions	A0711	Interrogation Problem		
									A1107	Data Problem		
									A23	Use of Device Problem		
							F26	No Health Consequences or Impact	A0711	Interrogation Problem		
									A1107	Data Problem		
									A23	Use of Device Problem		
18130	2026	All other reportable incidents	60789	Intracardiac pacemaker			E2403	No Clinical Signs, Symptoms or Conditions	A0709	Device Sensing Problem		
									A070910	Undersensing		
							F26	No Health Consequences or Impact	A0709	Device Sensing Problem		
									A070910	Undersensing		
18131	2026	All other reportable incidents	60956	Pacing/defibrillation lead tunneller	CT1684	Cardiac pacemakers and associated devices	E210401	Shock from Patient Lead(s)	A071302	Inappropriate/Inadequate Shock/Stimulation		
							F08	Hospitalization or Prolonged Hospitalization	A071302	Inappropriate/Inadequate Shock/Stimulation		
							F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A071302	Inappropriate/Inadequate Shock/Stimulation		
18132	2026	All other reportable incidents	47265	Dual-chamber implantable pacemaker, rate-responsive			E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
18133	2026	All other reportable incidents	47265	Dual-chamber implantable pacemaker, rate-responsive			E2403	No Clinical Signs, Symptoms or Conditions	A0908	Incorrect, Inadequate or Imprecise Result or Readings		
							F26	No Health Consequences or Impact	A0908	Incorrect, Inadequate or Imprecise Result or Readings		
18134	2026	All other reportable incidents	47265	Dual-chamber implantable pacemaker, rate-responsive			E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
18135	2026	All other reportable incidents	47265	Dual-chamber implantable pacemaker, rate-responsive			E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
18136	2026	All other reportable incidents	35852	Single-chamber implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E210401	Shock from Patient Lead(s)	A071302	Inappropriate/Inadequate Shock/Stimulation		
							F08	Hospitalization or Prolonged Hospitalization	A071302	Inappropriate/Inadequate Shock/Stimulation		
							F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A071302	Inappropriate/Inadequate Shock/Stimulation		
18137	2026	All other reportable incidents	47265	Dual-chamber implantable pacemaker, rate-responsive			E2403	No Clinical Signs, Symptoms or Conditions	A0908	Incorrect, Inadequate or Imprecise Result or Readings		
							F26	No Health Consequences or Impact	A0908	Incorrect, Inadequate or Imprecise Result or Readings		
18138	2026	All other reportable incidents	47265	Dual-chamber implantable pacemaker, rate-responsive			E2403	No Clinical Signs, Symptoms or Conditions	A0711	Interrogation Problem		
									A1107	Data Problem		
									A23	Use of Device Problem		
							F26	No Health Consequences or Impact	A0711	Interrogation Problem		
									A1107	Data Problem		
									A23	Use of Device Problem		
18139	2026	All other reportable incidents	47265	Dual-chamber implantable pacemaker, rate-responsive			E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	01/05/2026	01/05/2026
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	01/05/2026	01/05/2026
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	01/05/2026	01/05/2026
18140	2026	All other reportable incidents	47265	Dual-chamber implantable pacemaker, rate-responsive			E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
18141	2026	All other reportable incidents	47265	Dual-chamber implantable pacemaker, rate-responsive			E1906	Unspecified Infection	A24	No Apparent Device Problem		
							F08	Hospitalization or Prolonged Hospitalization	A24	No Apparent Device Problem		
							F1903	Device Explantation	A24	No Apparent Device Problem		
							F2303	Medication Required	A24	No Apparent Device Problem		
18142	2026	All other reportable incidents	47265	Dual-chamber implantable pacemaker, rate-responsive			E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	29/05/2025	29/05/2025
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	29/05/2025	29/05/2025
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	29/05/2025	29/05/2025
18143	2026	All other reportable incidents	47265	Dual-chamber implantable pacemaker, rate-responsive	CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	16/02/2026	16/02/2026
									A071102	Failure to Interrogate	16/02/2026	16/02/2026
							F1901	Additional Surgery	A070504	Premature Discharge of Battery	16/02/2026	16/02/2026
									A071102	Failure to Interrogate	16/02/2026	16/02/2026
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	16/02/2026	16/02/2026
									A071102	Failure to Interrogate	16/02/2026	16/02/2026
							F1901	Additional Surgery	A070504	Premature Discharge of Battery	16/02/2026	16/02/2026
									A071102	Failure to Interrogate	16/02/2026	16/02/2026
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	16/02/2026	16/02/2026
									A071102	Failure to Interrogate	16/02/2026	16/02/2026
							F1901	Additional Surgery	A070504	Premature Discharge of Battery	16/02/2026	16/02/2026
									A071102	Failure to Interrogate	16/02/2026	16/02/2026
18144	2026	All other reportable incidents	47265	Dual-chamber implantable pacemaker, rate-responsive			E020201	Anxiety	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
18145	2026	All other reportable incidents	47265	Dual-chamber implantable pacemaker, rate-responsive			E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
18146	2026		47265	Dual-chamber implantable pacemaker, rate-responsive	CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A27	Appropriate Device Problem Term/Code not Available		

							F1905	Device Revision or Replacement	A27	Appropriate Device Problem Term/Code not Available			
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A27	Appropriate Device Problem Term/Code not Available			
							F1905	Device Revision or Replacement	A27	Appropriate Device Problem Term/Code not Available			
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A27	Appropriate Device Problem Term/Code not Available			
							F1905	Device Revision or Replacement	A27	Appropriate Device Problem Term/Code not Available			
18147	2026	All other reportable incidents	47265	Dual-chamber implantable pacemaker, rate-responsive	CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem			
									A070504	Premature Discharge of Battery			
							F1905	Device Revision or Replacement	A0705	Battery Problem			
									A070504	Premature Discharge of Battery			
							F26	No Health Consequences or Impact	A0705	Battery Problem			
									A070504	Premature Discharge of Battery			
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem			
									A070504	Premature Discharge of Battery			
							F1905	Device Revision or Replacement	A0705	Battery Problem			
									A070504	Premature Discharge of Battery			
							F26	No Health Consequences or Impact	A0705	Battery Problem			
									A070504	Premature Discharge of Battery			
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem			
									A070504	Premature Discharge of Battery			
							F1905	Device Revision or Replacement	A0705	Battery Problem			
									A070504	Premature Discharge of Battery			
							F26	No Health Consequences or Impact	A0705	Battery Problem			
									A070504	Premature Discharge of Battery			
18148	2026	All other reportable incidents	47265	Dual-chamber implantable pacemaker, rate-responsive	CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem			
									A070504	Premature Discharge of Battery			
							F26	No Health Consequences or Impact	A0705	Battery Problem			
									A070504	Premature Discharge of Battery			
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem			
									A070504	Premature Discharge of Battery			
							F26	No Health Consequences or Impact	A0705	Battery Problem			
									A070504	Premature Discharge of Battery			
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem			
									A070504	Premature Discharge of Battery			
							F26	No Health Consequences or Impact	A0705	Battery Problem			
									A070504	Premature Discharge of Battery			
18149	2026	All other reportable incidents	47804	Implantable cardiac monitor	CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A0401	Break	20/01/2026	20/01/2026	
									A0413	Material Separation	20/01/2026	20/01/2026	
									A0511	Structural Problem	20/01/2026	20/01/2026	
							F1908	Prolonged Surgery	A0401	Break	20/01/2026	20/01/2026	
									A0413	Material Separation	20/01/2026	20/01/2026	
									A0511	Structural Problem	20/01/2026	20/01/2026	
							F26	No Health Consequences or Impact	A0401	Break	20/01/2026	20/01/2026	
									A0413	Material Separation	20/01/2026	20/01/2026	
									A0511	Structural Problem	20/01/2026	20/01/2026	
18150	2026		47263	Cardiac resynchronization therapy implantable pacemaker	CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem			
							F1901	Additional Surgery	A0705	Battery Problem			
							F2301	Additional Device Required	A0705	Battery Problem			
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem			
							F1901	Additional Surgery	A0705	Battery Problem			
							F2301	Additional Device Required	A0705	Battery Problem			
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem			
							F1901	Additional Surgery	A0705	Battery Problem			
							F2301	Additional Device Required	A0705	Battery Problem			
18151	2026	All other reportable incidents	37265	Dual-chamber implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem	23/02/2026	23/02/2026	
									A070504	Premature Discharge of Battery	23/02/2026	23/02/2026	
							F08	Hospitalization or Prolonged Hospitalization	A0705	Battery Problem	23/02/2026	23/02/2026	
									A070504	Premature Discharge of Battery	23/02/2026	23/02/2026	
							F1905	Device Revision or Replacement	A0705	Battery Problem	23/02/2026	23/02/2026	
									A070504	Premature Discharge of Battery	23/02/2026	23/02/2026	
18152	2026	All other reportable incidents	47265	Dual-chamber implantable pacemaker, rate-responsive			E2403	No Clinical Signs, Symptoms or Conditions	A020102	Nonstandard Device	15/04/2025	15/04/2025	
							F24	Insufficient Health Impact Information	A020102	Nonstandard Device	15/04/2025	15/04/2025	
18153	2026	All other reportable incidents	47265	Dual-chamber implantable pacemaker, rate-responsive			E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery			
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery			
18154	2026	All other reportable incidents	47265	Dual-chamber implantable pacemaker, rate-responsive			E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery			
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery			
18155	2026	All other reportable incidents	47265	Dual-chamber implantable pacemaker, rate-responsive			E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery			
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery			
18156	2026	All other reportable incidents	47267	Single-chamber implantable pacemaker, rate-responsive			E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery			
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery			
18157	2026	All other reportable incidents	47265	Dual-chamber implantable pacemaker, rate-responsive			E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery			
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery			
18158	2026	All other reportable incidents	47265	Dual-chamber implantable pacemaker, rate-responsive	CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A0203	Defective Device	02/03/2026	02/03/2026	
									A070504	Premature Discharge of Battery	02/03/2026	02/03/2026	
							F08	Hospitalization or Prolonged Hospitalization	A0203	Defective Device	02/03/2026	02/03/2026	
									A070504	Premature Discharge of Battery	02/03/2026	02/03/2026	
							F1901	Additional Surgery	A0203	Defective Device	02/03/2026	02/03/2026	
									A070504	Premature Discharge of Battery	02/03/2026	02/03/2026	
							F1905	Device Revision or Replacement	A0203	Defective Device	02/03/2026	02/03/2026	
									A070504	Premature Discharge of Battery	02/03/2026	02/03/2026	
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A0203	Defective Device	02/03/2026	02/03/2026	
									A070504	Premature Discharge of Battery	02/03/2026	02/03/2026	
							F08	Hospitalization or Prolonged Hospitalization	A0203	Defective Device	02/03/2026	02/03/2026	
									A070504	Premature Discharge of Battery	02/03/2026	02/03/2026	
							F1901	Additional Surgery	A0203	Defective Device	02/03/2026	02/03/2026	

									A070504	Premature Discharge of Battery	02/03/2026	02/03/2026
							F1905	Device Revision or Replacement	A0203	Defective Device	02/03/2026	02/03/2026
									A070504	Premature Discharge of Battery	02/03/2026	02/03/2026
					CT2750	Implantable cardiac electrical stimulators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A0203	Defective Device	02/03/2026	02/03/2026
									A070504	Premature Discharge of Battery	02/03/2026	02/03/2026
							F08	Hospitalization or Prolonged Hospitalization	A0203	Defective Device	02/03/2026	02/03/2026
									A070504	Premature Discharge of Battery	02/03/2026	02/03/2026
							F1901	Additional Surgery	A0203	Defective Device	02/03/2026	02/03/2026
									A070504	Premature Discharge of Battery	02/03/2026	02/03/2026
							F1905	Device Revision or Replacement	A0203	Defective Device	02/03/2026	02/03/2026
									A070504	Premature Discharge of Battery	02/03/2026	02/03/2026
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A0203	Defective Device	02/03/2026	02/03/2026
									A070504	Premature Discharge of Battery	02/03/2026	02/03/2026
							F08	Hospitalization or Prolonged Hospitalization	A0203	Defective Device	02/03/2026	02/03/2026
									A070504	Premature Discharge of Battery	02/03/2026	02/03/2026
							F1901	Additional Surgery	A0203	Defective Device	02/03/2026	02/03/2026
									A070504	Premature Discharge of Battery	02/03/2026	02/03/2026
							F1905	Device Revision or Replacement	A0203	Defective Device	02/03/2026	02/03/2026
									A070504	Premature Discharge of Battery	02/03/2026	02/03/2026
18159	2026	All other reportable incidents	47265	Dual-chamber implantable pacemaker, rate-responsive			E2403	No Clinical Signs, Symptoms or Conditions	A071102	Failure to Interrogate		
							F26	No Health Consequences or Impact	A071102	Failure to Interrogate		
18160	2026	All other reportable incidents	47267	Single-chamber implantable pacemaker, rate-responsive			E020201	Anxiety	A26	Insufficient Device Problem Information		
							F11	Non-Serious Injury/ Illness/ Impairment	A26	Insufficient Device Problem Information		
18161	2026	All other reportable incidents	47265	Dual-chamber implantable pacemaker, rate-responsive			E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
18162	2026	All other reportable incidents	47265	Dual-chamber implantable pacemaker, rate-responsive			E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
18163	2026	All other reportable incidents	47265	Dual-chamber implantable pacemaker, rate-responsive			E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
18164	2026	All other reportable incidents	47265	Dual-chamber implantable pacemaker, rate-responsive			E2403	No Clinical Signs, Symptoms or Conditions	A070909	Oversensing		
									A23	Use of Device Problem		
							F26	No Health Consequences or Impact	A070909	Oversensing		
									A23	Use of Device Problem		
18165	2026	All other reportable incidents	47265	Dual-chamber implantable pacemaker, rate-responsive	CT1685	Cardiac pacemakers	E010701	Confusion/ Disorientation	A0908	Incorrect, Inadequate or Imprecise Result or Readings	25/02/2026	25/02/2026
							E020201	Anxiety	A0908	Incorrect, Inadequate or Imprecise Result or Readings	25/02/2026	25/02/2026
							E020204	Malaise	A0908	Incorrect, Inadequate or Imprecise Result or Readings	25/02/2026	25/02/2026
							F0103	Unexpected Therapeutic Effects	A0908	Incorrect, Inadequate or Imprecise Result or Readings	25/02/2026	25/02/2026
							F08	Hospitalization or Prolonged Hospitalization	A0908	Incorrect, Inadequate or Imprecise Result or Readings	25/02/2026	25/02/2026
							F1905	Device Revision or Replacement	A0908	Incorrect, Inadequate or Imprecise Result or Readings	25/02/2026	25/02/2026
					CT257	Implantable cardiac pacemakers	E010701	Confusion/ Disorientation	A0908	Incorrect, Inadequate or Imprecise Result or Readings	25/02/2026	25/02/2026
							E020201	Anxiety	A0908	Incorrect, Inadequate or Imprecise Result or Readings	25/02/2026	25/02/2026
							E020204	Malaise	A0908	Incorrect, Inadequate or Imprecise Result or Readings	25/02/2026	25/02/2026
							F0103	Unexpected Therapeutic Effects	A0908	Incorrect, Inadequate or Imprecise Result or Readings	25/02/2026	25/02/2026
							F08	Hospitalization or Prolonged Hospitalization	A0908	Incorrect, Inadequate or Imprecise Result or Readings	25/02/2026	25/02/2026
							F1905	Device Revision or Replacement	A0908	Incorrect, Inadequate or Imprecise Result or Readings	25/02/2026	25/02/2026
					CT2866	Implantable cardiac electrical stimulators	E010701	Confusion/ Disorientation	A0908	Incorrect, Inadequate or Imprecise Result or Readings	25/02/2026	25/02/2026
							E020201	Anxiety	A0908	Incorrect, Inadequate or Imprecise Result or Readings	25/02/2026	25/02/2026
							E020204	Malaise	A0908	Incorrect, Inadequate or Imprecise Result or Readings	25/02/2026	25/02/2026
							F0103	Unexpected Therapeutic Effects	A0908	Incorrect, Inadequate or Imprecise Result or Readings	25/02/2026	25/02/2026
							F08	Hospitalization or Prolonged Hospitalization	A0908	Incorrect, Inadequate or Imprecise Result or Readings	25/02/2026	25/02/2026
							F1905	Device Revision or Replacement	A0908	Incorrect, Inadequate or Imprecise Result or Readings	25/02/2026	25/02/2026
18166	2026	All other reportable incidents	47265	Dual-chamber implantable pacemaker, rate-responsive			E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
18167	2026	All other reportable incidents	47267	Single-chamber implantable pacemaker, rate-responsive			E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
18168	2026	All other reportable incidents	47265	Dual-chamber implantable pacemaker, rate-responsive			E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
18169	2026	All other reportable incidents	47265	Dual-chamber implantable pacemaker, rate-responsive			E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
18170	2026	All other reportable incidents	47265	Dual-chamber implantable pacemaker, rate-responsive			E2403	No Clinical Signs, Symptoms or Conditions	A0908	Incorrect, Inadequate or Imprecise Result or Readings		
							F26	No Health Consequences or Impact	A0908	Incorrect, Inadequate or Imprecise Result or Readings		
18171	2026	All other reportable incidents	47265	Dual-chamber implantable pacemaker, rate-responsive			E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
18172	2026	All other reportable incidents	47265	Dual-chamber implantable pacemaker, rate-responsive			E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	12/03/2026	12/03/2026
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	12/03/2026	12/03/2026
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	12/03/2026	12/03/2026
18173	2026	All other reportable incidents	47265	Dual-chamber implantable pacemaker, rate-responsive			E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	02/06/2026	02/06/2026
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	02/06/2026	02/06/2026
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	02/06/2026	02/06/2026
18174	2026	All other reportable incidents	47265	Dual-chamber implantable pacemaker, rate-responsive			E2403	No Clinical Signs, Symptoms or Conditions	A020102	Nonstandard Device	01/05/2025	01/05/2025
							F24	Insufficient Health Impact Information	A020102	Nonstandard Device	01/05/2025	01/05/2025
18175	2026	All other reportable incidents	47263	Cardiac resynchronization therapy implantable	CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A070501	Battery Problem: High Impedance		

				pacemaker					A070504	Premature Discharge of Battery		
									A090405	Output Below Specifications		
							F1903	Device Explantation	A070501	Battery Problem: High Impedance		
									A070504	Premature Discharge of Battery		
									A090405	Output Below Specifications		
							F26	No Health Consequences or Impact	A070501	Battery Problem: High Impedance		
									A070504	Premature Discharge of Battery		
									A090405	Output Below Specifications		
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A070501	Battery Problem: High Impedance		
									A070504	Premature Discharge of Battery		
									A090405	Output Below Specifications		
							F1903	Device Explantation	A070501	Battery Problem: High Impedance		
									A070504	Premature Discharge of Battery		
									A090405	Output Below Specifications		
							F26	No Health Consequences or Impact	A070501	Battery Problem: High Impedance		
									A070504	Premature Discharge of Battery		
									A090405	Output Below Specifications		
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A070501	Battery Problem: High Impedance		
									A070504	Premature Discharge of Battery		
									A090405	Output Below Specifications		
							F1903	Device Explantation	A070501	Battery Problem: High Impedance		
									A070504	Premature Discharge of Battery		
									A090405	Output Below Specifications		
							F26	No Health Consequences or Impact	A070501	Battery Problem: High Impedance		
									A070504	Premature Discharge of Battery		
									A090405	Output Below Specifications		
18176	2026	All other reportable incidents	47265	Dual-chamber implantable pacemaker, rate-responsive			E2403	No Clinical Signs, Symptoms or Conditions	A0203	Defective Device		
									A070504	Premature Discharge of Battery		
							F1905	Device Revision or Replacement	A0203	Defective Device		
									A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A0203	Defective Device		
									A070504	Premature Discharge of Battery		
18177	2026	All other reportable incidents	60789	Intracardiac pacemaker			E2403	No Clinical Signs, Symptoms or Conditions	A0701	Capturing Problem		
									A070102	High Capture Threshold		
									A0712	Pacing Problem		
							F26	No Health Consequences or Impact	A0701	Capturing Problem		
									A070102	High Capture Threshold		
									A0712	Pacing Problem		
18178	2026	All other reportable incidents	47265	Dual-chamber implantable pacemaker, rate-responsive			E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	24/04/2026	24/04/2026
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	24/04/2026	24/04/2026
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	24/04/2026	24/04/2026
18179	2026	All other reportable incidents	47265	Dual-chamber implantable pacemaker, rate-responsive			E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
18180	2026	All other reportable incidents	47265	Dual-chamber implantable pacemaker, rate-responsive			E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
18181	2026	All other reportable incidents	47265	Dual-chamber implantable pacemaker, rate-responsive			E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
18182	2026	All other reportable incidents	47267	Single-chamber implantable pacemaker, rate-responsive			E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
18183	2026	All other reportable incidents	47265	Dual-chamber implantable pacemaker, rate-responsive			E233001	Chest Pain	A070909	Oversensing		
							F26	No Health Consequences or Impact	A070909	Oversensing		
18184	2026	All other reportable incidents	47265	Dual-chamber implantable pacemaker, rate-responsive			E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
18185	2026	All other reportable incidents	47265	Dual-chamber implantable pacemaker, rate-responsive			E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
18186	2026	All other reportable incidents	47265	Dual-chamber implantable pacemaker, rate-responsive			E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
18187	2026	All other reportable incidents	47265	Dual-chamber implantable pacemaker, rate-responsive			E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
18188	2026	All other reportable incidents	47265	Dual-chamber implantable pacemaker, rate-responsive			E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	08/05/2026	08/05/2026
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	08/05/2026	08/05/2026
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	08/05/2026	08/05/2026
18189	2026	All other reportable incidents	47265	Dual-chamber implantable pacemaker, rate-responsive			E2403	No Clinical Signs, Symptoms or Conditions	A0712	Pacing Problem		
							F26	No Health Consequences or Impact	A0712	Pacing Problem		
18190	2026	All other reportable incidents	47265	Dual-chamber implantable pacemaker, rate-responsive			E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem		
							F26	No Health Consequences or Impact	A0705	Battery Problem		
18191	2026	All other reportable incidents	47270	Cardiac resynchronization therapy implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem		
									A160601	Premature Elective Replacement Indicator		
							F1901	Additional Surgery	A0705	Battery Problem		
									A160601	Premature Elective Replacement Indicator		
							F1905	Device Revision or Replacement	A0705	Battery Problem		
									A160601	Premature Elective Replacement Indicator		
							F26	No Health Consequences or Impact	A0705	Battery Problem		
									A160601	Premature Elective Replacement Indicator		
18192	2026	All other reportable incidents	47270	Cardiac resynchronization therapy implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E210401	Shock from Patient Lead(s)	A070909	Oversensing		
									A071302	Inappropriate/Inadequate Shock/Stimulation		
									A090801	Signal Artifact/Noise		
									A1301	Failure to Read Input Signal		
									A23	Use of Device Problem		
							F2203	Imaging Required	A070909	Oversensing		
									A071302	Inappropriate/Inadequate Shock/Stimulation		
									A090801	Signal Artifact/Noise		
									A1301	Failure to Read Input Signal		
									A23	Use of Device Problem		
							F1009	Inappropriate Treatment	A070909	Oversensing		
									A071302	Inappropriate/Inadequate Shock/Stimulation		
									A090801	Signal Artifact/Noise		
									A1301	Failure to Read Input Signal		

									A23	Use of Device Problem			
18193	2026	All other reportable incidents	47265	Dual-chamber implantable pacemaker, rate-responsive			E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery			
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery			
18194	2026	All other reportable incidents	60956	Pacing/defibrillation lead tunneller	CT1684	Cardiac pacemakers and associated devices	E210401	Shock from Patient Lead(s)	A070909	Oversensing			
									A071302	Inappropriate/Inadequate Shock/Stimulation			
									A090801	Signal Artifact/Noise			
									A1301	Failure to Read Input Signal			
							F2203	Imaging Required	A070909	Oversensing			
									A071302	Inappropriate/Inadequate Shock/Stimulation			
									A090801	Signal Artifact/Noise			
									A1301	Failure to Read Input Signal			
							F1009	Inappropriate Treatment	A070909	Oversensing			
									A071302	Inappropriate/Inadequate Shock/Stimulation			
									A090801	Signal Artifact/Noise			
									A1301	Failure to Read Input Signal			
18195	2026	All other reportable incidents	47265	Dual-chamber implantable pacemaker, rate-responsive			E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery			
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery			
18196	2026	All other reportable incidents	47265	Dual-chamber implantable pacemaker, rate-responsive			E2403	No Clinical Signs, Symptoms or Conditions	A071102	Failure to Interrogate			
							F26	No Health Consequences or Impact	A071102	Failure to Interrogate			
18197	2026	All other reportable incidents	47265	Dual-chamber implantable pacemaker, rate-responsive			E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery			
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery			
18198	2026	All other reportable incidents	47265	Dual-chamber implantable pacemaker, rate-responsive	CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	04/11/2025	04/11/2025	
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	04/11/2025	04/11/2025	
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	04/11/2025	04/11/2025	
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	04/11/2025	04/11/2025	
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	04/11/2025	04/11/2025	
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	04/11/2025	04/11/2025	
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	04/11/2025	04/11/2025	
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	04/11/2025	04/11/2025	
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	04/11/2025	04/11/2025	
18199	2026	All other reportable incidents	47265	Dual-chamber implantable pacemaker, rate-responsive			E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery			
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery			
18200	2026	All other reportable incidents	17882	Physiologic-monitoring defibrillation system	CT1684	Cardiac pacemakers and associated devices	E0602	Cardiac Arrest	A0203	Defective Device			
							F05	Delay to Treatment/ Therapy	A0203	Defective Device			
18201	2026	All other reportable incidents	47265	Dual-chamber implantable pacemaker, rate-responsive			E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	16/03/2026	16/03/2026	
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	16/03/2026	16/03/2026	
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	16/03/2026	16/03/2026	
18202	2026	All other reportable incidents	47265	Dual-chamber implantable pacemaker, rate-responsive			E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery			
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery			
18203	2026	All other reportable incidents	47265	Dual-chamber implantable pacemaker, rate-responsive			E060104	Bradycardia	A070909	Oversensing			
									A0712	Pacing Problem			
							F1001	Absence of Treatment	A070909	Oversensing			
									A0712	Pacing Problem			
18204	2026	All other reportable incidents	47265	Dual-chamber implantable pacemaker, rate-responsive			E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery			
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery			
18205	2026	All other reportable incidents	47265	Dual-chamber implantable pacemaker, rate-responsive			E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery			
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery			
18206	2026	All other reportable incidents	47265	Dual-chamber implantable pacemaker, rate-responsive			E2403	No Clinical Signs, Symptoms or Conditions	A1605	Reset Problem			
									A160503	Inappropriate or Unexpected Reset			
							F26	No Health Consequences or Impact	A1605	Reset Problem			
									A160503	Inappropriate or Unexpected Reset			
18207	2026		47265	Dual-chamber implantable pacemaker, rate-responsive	CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem			
									A070502	Battery Problem: Low Impedance			
							F1905	Device Revision or Replacement	A0705	Battery Problem			
									A070502	Battery Problem: Low Impedance			
							F26	No Health Consequences or Impact	A0705	Battery Problem			
									A070502	Battery Problem: Low Impedance			
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem			
									A070502	Battery Problem: Low Impedance			
							F1905	Device Revision or Replacement	A0705	Battery Problem			
									A070502	Battery Problem: Low Impedance			
							F26	No Health Consequences or Impact	A0705	Battery Problem			
									A070502	Battery Problem: Low Impedance			
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem			
									A070502	Battery Problem: Low Impedance			
							F1905	Device Revision or Replacement	A0705	Battery Problem			
									A070502	Battery Problem: Low Impedance			
							F26	No Health Consequences or Impact	A0705	Battery Problem			
									A070502	Battery Problem: Low Impedance			
18208	2026	All other reportable incidents	47265	Dual-chamber implantable pacemaker, rate-responsive	CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem			
									A070504	Premature Discharge of Battery			
							F26	No Health Consequences or Impact	A0705	Battery Problem			
									A070504	Premature Discharge of Battery			
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem			
									A070504	Premature Discharge of Battery			
							F26	No Health Consequences or Impact	A0705	Battery Problem			
									A070504	Premature Discharge of Battery			
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem			
									A070504	Premature Discharge of Battery			
							F26	No Health Consequences or Impact	A0705	Battery Problem			
									A070504	Premature Discharge of Battery			
18209	2026		47265	Dual-chamber implantable pacemaker, rate-responsive	CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A0201	Product Quality Problem			
									A0705	Battery Problem			
							F26	No Health Consequences or Impact	A0201	Product Quality Problem			
									A0705	Battery Problem			
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A0201	Product Quality Problem			
									A0705	Battery Problem			
							F26	No Health Consequences or Impact	A0201	Product Quality Problem			

									A0705	Battery Problem		
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A0201	Product Quality Problem		
									A0705	Battery Problem		
							F26	No Health Consequences or Impact	A0201	Product Quality Problem		
									A0705	Battery Problem		
18210	2026	All other reportable incidents	47265	Dual-chamber implantable pacemaker, rate-responsive			E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
18211	2026	All other reportable incidents	47265	Dual-chamber implantable pacemaker, rate-responsive			E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
18212	2026		60956	Pacing/defibrillation lead tunneller	CT1684	Cardiac pacemakers and associated devices	E210401	Shock from Patient Lead(s)	A0104	Migration or Expulsion of Device		
									A071302	Inappropriate/Inadequate Shock/Stimulation		
							F1903	Device Explantation	A0104	Migration or Expulsion of Device		
									A071302	Inappropriate/Inadequate Shock/Stimulation		
							F2307	Reprogramming of Device	A0104	Migration or Expulsion of Device		
									A071302	Inappropriate/Inadequate Shock/Stimulation		
							F25	Unanticipated Adverse Device Effect	A0104	Migration or Expulsion of Device		
									A071302	Inappropriate/Inadequate Shock/Stimulation		
18213	2026		47263	Cardiac resynchronization therapy implantable pacemaker	CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A0202	Defective Component		
									A1103	Program or Algorithm Execution Problem		
							F1905	Device Revision or Replacement	A0202	Defective Component		
									A1103	Program or Algorithm Execution Problem		
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A0202	Defective Component		
									A1103	Program or Algorithm Execution Problem		
							F1905	Device Revision or Replacement	A0202	Defective Component		
									A1103	Program or Algorithm Execution Problem		
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A0202	Defective Component		
									A1103	Program or Algorithm Execution Problem		
							F1905	Device Revision or Replacement	A0202	Defective Component		
									A1103	Program or Algorithm Execution Problem		
18214	2026	All other reportable incidents	47265	Dual-chamber implantable pacemaker, rate-responsive			E2403	No Clinical Signs, Symptoms or Conditions	A0712	Pacing Problem		
									A072201	High Impedance		
							F15	Recognized Device or Procedural Complication	A0712	Pacing Problem		
									A072201	High Impedance		
18215	2026	All other reportable incidents	47265	Dual-chamber implantable pacemaker, rate-responsive			E0601	Arrhythmia	A070504	Premature Discharge of Battery		
									A071102	Failure to Interrogate		
							E2311	Discomfort	A070504	Premature Discharge of Battery		
									A071102	Failure to Interrogate		
							E2312	Fatigue	A070504	Premature Discharge of Battery		
									A071102	Failure to Interrogate		
							F11	Non-Serious Injury/ Illness/ Impairment	A070504	Premature Discharge of Battery		
									A071102	Failure to Interrogate		
18216	2026		47265	Dual-chamber implantable pacemaker, rate-responsive	CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem		
							F1905	Device Revision or Replacement	A0705	Battery Problem		
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem		
							F1905	Device Revision or Replacement	A0705	Battery Problem		
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem		
							F1905	Device Revision or Replacement	A0705	Battery Problem		
18217	2026	All other reportable incidents	47265	Dual-chamber implantable pacemaker, rate-responsive	CT1685	Cardiac pacemakers	E012207	Twitching	A0705	Battery Problem	13/03/2026	13/03/2026
									A0908	Incorrect, Inadequate or Imprecise Result or Readings	13/03/2026	13/03/2026
							F08	Hospitalization or Prolonged Hospitalization	A0705	Battery Problem	13/03/2026	13/03/2026
									A0908	Incorrect, Inadequate or Imprecise Result or Readings	13/03/2026	13/03/2026
							F1905	Device Revision or Replacement	A0705	Battery Problem	13/03/2026	13/03/2026
									A0908	Incorrect, Inadequate or Imprecise Result or Readings	13/03/2026	13/03/2026
					CT257	Implantable cardiac pacemakers	E012207	Twitching	A0705	Battery Problem	13/03/2026	13/03/2026
									A0908	Incorrect, Inadequate or Imprecise Result or Readings	13/03/2026	13/03/2026
							F08	Hospitalization or Prolonged Hospitalization	A0705	Battery Problem	13/03/2026	13/03/2026
									A0908	Incorrect, Inadequate or Imprecise Result or Readings	13/03/2026	13/03/2026
							F1905	Device Revision or Replacement	A0705	Battery Problem	13/03/2026	13/03/2026
									A0908	Incorrect, Inadequate or Imprecise Result or Readings	13/03/2026	13/03/2026
					CT2866	Implantable cardiac electrical stimulators	E012207	Twitching	A0705	Battery Problem	13/03/2026	13/03/2026
									A0908	Incorrect, Inadequate or Imprecise Result or Readings	13/03/2026	13/03/2026
							F08	Hospitalization or Prolonged Hospitalization	A0705	Battery Problem	13/03/2026	13/03/2026
									A0908	Incorrect, Inadequate or Imprecise Result or Readings	13/03/2026	13/03/2026
							F1905	Device Revision or Replacement	A0705	Battery Problem	13/03/2026	13/03/2026
									A0908	Incorrect, Inadequate or Imprecise Result or Readings	13/03/2026	13/03/2026
18218	2026		47265	Dual-chamber implantable pacemaker, rate-responsive	CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A071205	Pacemaker Found in Back-Up Mode		
							F01	Change in Therapeutic Response	A071205	Pacemaker Found in Back-Up Mode		
							F26	No Health Consequences or Impact	A071205	Pacemaker Found in Back-Up Mode		
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A071205	Pacemaker Found in Back-Up Mode		
							F01	Change in Therapeutic Response	A071205	Pacemaker Found in Back-Up Mode		
							F26	No Health Consequences or Impact	A071205	Pacemaker Found in Back-Up Mode		
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A071205	Pacemaker Found in Back-Up Mode		
							F01	Change in Therapeutic Response	A071205	Pacemaker Found in Back-Up Mode		
							F26	No Health Consequences or Impact	A071205	Pacemaker Found in Back-Up Mode		
18219	2026	All other reportable incidents	47265	Dual-chamber implantable pacemaker, rate-responsive			E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
18220	2026	All other reportable incidents	47265	Dual-chamber implantable pacemaker, rate-responsive			E2403	No Clinical Signs, Symptoms or Conditions	A0709	Device Sensing Problem		
									A070909	Oversensing		
							F26	No Health Consequences or Impact	A0709	Device Sensing Problem		
									A070909	Oversensing		
18221	2026		47265	Dual-chamber implantable pacemaker, rate-	CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A1108	Date/Time-Related Software Problem		

				responsive			F26	No Health Consequences or Impact	A1108	Date/Time-Related Software Problem			
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A1108	Date/Time-Related Software Problem			
							F26	No Health Consequences or Impact	A1108	Date/Time-Related Software Problem			
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A1108	Date/Time-Related Software Problem			
							F26	No Health Consequences or Impact	A1108	Date/Time-Related Software Problem			
18222	2026	All other reportable incidents	47265	Dual-chamber implantable pacemaker, rate-responsive			E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery			
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery			
18223	2026	All other reportable incidents	47265	Dual-chamber implantable pacemaker, rate-responsive			E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	06/05/2026	06/05/2026	
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	06/05/2026	06/05/2026	
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	06/05/2026	06/05/2026	
18224	2026	All other reportable incidents	47265	Dual-chamber implantable pacemaker, rate-responsive			E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery			
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery			
18225	2026	All other reportable incidents	47265	Dual-chamber implantable pacemaker, rate-responsive			E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery			
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery			
18226	2026	All other reportable incidents	47267	Single-chamber implantable pacemaker, rate-responsive			E2403	No Clinical Signs, Symptoms or Conditions	A071102	Failure to Interrogate			
									A13	Communication or Transmission Problem			
							F26	No Health Consequences or Impact	A071102	Failure to Interrogate			
									A13	Communication or Transmission Problem			
18227	2026	All other reportable incidents	47265	Dual-chamber implantable pacemaker, rate-responsive	CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A0203	Defective Device			
									A0705	Battery Problem			
									A070504	Premature Discharge of Battery			
							F12	Serious Injury/ Illness/ Impairment	A0203	Defective Device			
									A0705	Battery Problem			
									A070504	Premature Discharge of Battery			
							F26	No Health Consequences or Impact	A0203	Defective Device			
									A0705	Battery Problem			
									A070504	Premature Discharge of Battery			
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A0203	Defective Device			
									A0705	Battery Problem			
									A070504	Premature Discharge of Battery			
							F12	Serious Injury/ Illness/ Impairment	A0203	Defective Device			
									A0705	Battery Problem			
									A070504	Premature Discharge of Battery			
							F26	No Health Consequences or Impact	A0203	Defective Device			
									A0705	Battery Problem			
									A070504	Premature Discharge of Battery			
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A0203	Defective Device			
									A0705	Battery Problem			
									A070504	Premature Discharge of Battery			
							F12	Serious Injury/ Illness/ Impairment	A0203	Defective Device			
									A0705	Battery Problem			
									A070504	Premature Discharge of Battery			
							F26	No Health Consequences or Impact	A0203	Defective Device			
									A0705	Battery Problem			
									A070504	Premature Discharge of Battery			
18228	2026	All other reportable incidents	47265	Dual-chamber implantable pacemaker, rate-responsive			E0601	Arrhythmia	A070101	Failure to Capture	12/03/2026	12/03/2026	
									A070504	Premature Discharge of Battery	12/03/2026	12/03/2026	
									A072202	Low Impedance	12/03/2026	12/03/2026	
							F1901	Additional Surgery	A070101	Failure to Capture	12/03/2026	12/03/2026	
									A070504	Premature Discharge of Battery	12/03/2026	12/03/2026	
									A072202	Low Impedance	12/03/2026	12/03/2026	
18229	2026	All other reportable incidents	47265	Dual-chamber implantable pacemaker, rate-responsive			E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery			
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery			
18230	2026	All other reportable incidents	47265	Dual-chamber implantable pacemaker, rate-responsive			E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery			
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery			
18231	2026	All other reportable incidents	47265	Dual-chamber implantable pacemaker, rate-responsive			E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery			
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery			
18232	2026	All other reportable incidents	47267	Single-chamber implantable pacemaker, rate-responsive			E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery			
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery			
18233	2026	All other reportable incidents	47265	Dual-chamber implantable pacemaker, rate-responsive			E0119	Loss of Consciousness	A070101	Failure to Capture			
									A070102	High Capture Threshold			
									A0712	Pacing Problem			
							E011903	Syncope/Fainting	A070101	Failure to Capture			
									A070102	High Capture Threshold			
									A0712	Pacing Problem			
							F1001	Absence of Treatment	A070101	Failure to Capture			
									A070102	High Capture Threshold			
									A0712	Pacing Problem			
18234	2026	All other reportable incidents	47267	Single-chamber implantable pacemaker, rate-responsive			E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	10/12/2025	10/12/2025	
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	10/12/2025	10/12/2025	
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	10/12/2025	10/12/2025	
18235	2026	All other reportable incidents	47265	Dual-chamber implantable pacemaker, rate-responsive			E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery			
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery			
18236	2026	All other reportable incidents	47265	Dual-chamber implantable pacemaker, rate-responsive			E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem			
									A070504	Premature Discharge of Battery			
							F26	No Health Consequences or Impact	A0705	Battery Problem			
									A070504	Premature Discharge of Battery			
18237	2026	All other reportable incidents	47265	Dual-chamber implantable pacemaker, rate-responsive			E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	17/04/2026	17/04/2026	
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	17/04/2026	17/04/2026	
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	17/04/2026	17/04/2026	
18238	2026	All other reportable incidents	47265	Dual-chamber implantable pacemaker, rate-responsive			E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery			
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery			
18239	2026	All other reportable incidents	47265	Dual-chamber implantable pacemaker, rate-responsive			E0717	Dyspnea	A070504	Premature Discharge of Battery			
							E2320	High Blood Pressure/ Hypertension	A070504	Premature Discharge of Battery			
							F11	Non-Serious Injury/ Illness/ Impairment	A070504	Premature Discharge of Battery			
18240	2026	All other reportable incidents	47267	Single-chamber implantable pacemaker, rate-responsive			E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery			
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery			
18241	2026		47263	Cardiac resynchronization therapy implantable	CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A2205	Device Provides Uninformative Message			

				pacemaker					A27	Appropriate Device Problem Term/Code not Available		
							F26	No Health Consequences or Impact	A2205	Device Provides Uninformative Message		
									A27	Appropriate Device Problem Term/Code not Available		
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A2205	Device Provides Uninformative Message		
									A27	Appropriate Device Problem Term/Code not Available		
							F26	No Health Consequences or Impact	A2205	Device Provides Uninformative Message		
									A27	Appropriate Device Problem Term/Code not Available		
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A2205	Device Provides Uninformative Message		
									A27	Appropriate Device Problem Term/Code not Available		
							F26	No Health Consequences or Impact	A2205	Device Provides Uninformative Message		
									A27	Appropriate Device Problem Term/Code not Available		
18242	2026		47265	Dual-chamber implantable pacemaker, rate-responsive	CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A0203	Defective Device	24/03/2026	24/03/2026
									A0510	Retraction Problem	24/03/2026	24/03/2026
									A1204	Failure to Disconnect	24/03/2026	24/03/2026
							F1908	Prolonged Surgery	A0203	Defective Device	24/03/2026	24/03/2026
									A0510	Retraction Problem	24/03/2026	24/03/2026
									A1204	Failure to Disconnect	24/03/2026	24/03/2026
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A0203	Defective Device	24/03/2026	24/03/2026
									A0510	Retraction Problem	24/03/2026	24/03/2026
									A1204	Failure to Disconnect	24/03/2026	24/03/2026
							F1908	Prolonged Surgery	A0203	Defective Device	24/03/2026	24/03/2026
									A0510	Retraction Problem	24/03/2026	24/03/2026
									A1204	Failure to Disconnect	24/03/2026	24/03/2026
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A0203	Defective Device	24/03/2026	24/03/2026
									A0510	Retraction Problem	24/03/2026	24/03/2026
									A1204	Failure to Disconnect	24/03/2026	24/03/2026
							F1908	Prolonged Surgery	A0203	Defective Device	24/03/2026	24/03/2026
									A0510	Retraction Problem	24/03/2026	24/03/2026
									A1204	Failure to Disconnect	24/03/2026	24/03/2026
18243	2026	All other reportable incidents	47265	Dual-chamber implantable pacemaker, rate-responsive			E2403	No Clinical Signs, Symptoms or Conditions	A160601	Premature Elective Replacement Indicator	13/03/2026	13/03/2026
							F1903	Device Explantation	A160601	Premature Elective Replacement Indicator	13/03/2026	13/03/2026
18244	2026	All other reportable incidents	47265	Dual-chamber implantable pacemaker, rate-responsive	CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A070501	Battery Problem: High Impedance	26/03/2026	26/03/2026
									A0908	Incorrect, Inadequate or Imprecise Result or Readings	26/03/2026	26/03/2026
							F08	Hospitalization or Prolonged Hospitalization	A070501	Battery Problem: High Impedance	26/03/2026	26/03/2026
									A0908	Incorrect, Inadequate or Imprecise Result or Readings	26/03/2026	26/03/2026
							F1905	Device Revision or Replacement	A070501	Battery Problem: High Impedance	26/03/2026	26/03/2026
									A0908	Incorrect, Inadequate or Imprecise Result or Readings	26/03/2026	26/03/2026
							F26	No Health Consequences or Impact	A070501	Battery Problem: High Impedance	26/03/2026	26/03/2026
									A0908	Incorrect, Inadequate or Imprecise Result or Readings	26/03/2026	26/03/2026
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A070501	Battery Problem: High Impedance	26/03/2026	26/03/2026
									A0908	Incorrect, Inadequate or Imprecise Result or Readings	26/03/2026	26/03/2026
							F08	Hospitalization or Prolonged Hospitalization	A070501	Battery Problem: High Impedance	26/03/2026	26/03/2026
									A0908	Incorrect, Inadequate or Imprecise Result or Readings	26/03/2026	26/03/2026
							F1905	Device Revision or Replacement	A070501	Battery Problem: High Impedance	26/03/2026	26/03/2026
									A0908	Incorrect, Inadequate or Imprecise Result or Readings	26/03/2026	26/03/2026
							F26	No Health Consequences or Impact	A070501	Battery Problem: High Impedance	26/03/2026	26/03/2026
									A0908	Incorrect, Inadequate or Imprecise Result or Readings	26/03/2026	26/03/2026
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A070501	Battery Problem: High Impedance	26/03/2026	26/03/2026
									A0908	Incorrect, Inadequate or Imprecise Result or Readings	26/03/2026	26/03/2026
							F08	Hospitalization or Prolonged Hospitalization	A070501	Battery Problem: High Impedance	26/03/2026	26/03/2026
									A0908	Incorrect, Inadequate or Imprecise Result or Readings	26/03/2026	26/03/2026
							F1905	Device Revision or Replacement	A070501	Battery Problem: High Impedance	26/03/2026	26/03/2026
									A0908	Incorrect, Inadequate or Imprecise Result or Readings	26/03/2026	26/03/2026
							F26	No Health Consequences or Impact	A070501	Battery Problem: High Impedance	26/03/2026	26/03/2026
									A0908	Incorrect, Inadequate or Imprecise Result or Readings	26/03/2026	26/03/2026
18245	2026	All other reportable incidents	60789	Intracardiac pacemaker			E2403	No Clinical Signs, Symptoms or Conditions	A070101	Failure to Capture	20/03/2026	20/03/2026
									A1502	Positioning Problem	20/03/2026	20/03/2026
							F26	No Health Consequences or Impact	A070101	Failure to Capture	20/03/2026	20/03/2026
									A1502	Positioning Problem	20/03/2026	20/03/2026
18246	2026	All other reportable incidents	47270	Cardiac resynchronization therapy implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A0908	Incorrect, Inadequate or Imprecise Result or Readings		
									A1107	Data Problem		
							F26	No Health Consequences or Impact	A0908	Incorrect, Inadequate or Imprecise Result or Readings		
									A1107	Data Problem		
18247	2026	All other reportable incidents	47265	Dual-chamber implantable pacemaker, rate-responsive			E2403	No Clinical Signs, Symptoms or Conditions	A070909	Oversensing		
									A0712	Pacing Problem		
									A090801	Signal Artifact/Noise		
							F1001	Absence of Treatment	A070909	Oversensing		
									A0712	Pacing Problem		
									A090801	Signal Artifact/Noise		
18248	2026		47265	Dual-chamber implantable pacemaker, rate-responsive			E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	14/03/2026	14/03/2026
							F1901	Additional Surgery	A24	No Apparent Device Problem	14/03/2026	14/03/2026
18249	2026	All other reportable incidents	47265	Dual-chamber implantable pacemaker, rate-			E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		

				responsive			F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
18250	2026	All other reportable incidents	47265	Dual-chamber implantable pacemaker, rate-responsive			E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
18251	2026	All other reportable incidents	47267	Single-chamber implantable pacemaker, rate-responsive			E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
18252	2026	All other reportable incidents	47265	Dual-chamber implantable pacemaker, rate-responsive			E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
18253	2026	All other reportable incidents	47265	Dual-chamber implantable pacemaker, rate-responsive			E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
18254	2026	All other reportable incidents	47265	Dual-chamber implantable pacemaker, rate-responsive	CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	08/04/2026	08/04/2026
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	08/04/2026	08/04/2026
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	08/04/2026	08/04/2026
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	08/04/2026	08/04/2026
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	08/04/2026	08/04/2026
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	08/04/2026	08/04/2026
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	08/04/2026	08/04/2026
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	08/04/2026	08/04/2026
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	08/04/2026	08/04/2026
18255	2026	All other reportable incidents	47265	Dual-chamber implantable pacemaker, rate-responsive			E2403	No Clinical Signs, Symptoms or Conditions	A05	Mechanical Problem	23/03/2026	23/03/2026
							F26	No Health Consequences or Impact	A05	Mechanical Problem	23/03/2026	23/03/2026
18256	2026	All other reportable incidents	47265	Dual-chamber implantable pacemaker, rate-responsive			E2403	No Clinical Signs, Symptoms or Conditions	A071102	Failure to Interrogate		
							F26	No Health Consequences or Impact	A071102	Failure to Interrogate		
18257	2026	All other reportable incidents	47265	Dual-chamber implantable pacemaker, rate-responsive			E2403	No Clinical Signs, Symptoms or Conditions	A160105	False Alarm		
							F26	No Health Consequences or Impact	A160105	False Alarm		
18258	2026	All other reportable incidents	47265	Dual-chamber implantable pacemaker, rate-responsive			E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem		
							F12	Serious Injury/ Illness/ Impairment	A0705	Battery Problem		
							F23	Unexpected Medical Intervention	A0705	Battery Problem		
18259	2026	All other reportable incidents	47265	Dual-chamber implantable pacemaker, rate-responsive			E2403	No Clinical Signs, Symptoms or Conditions	A071102	Failure to Interrogate		
							F26	No Health Consequences or Impact	A071102	Failure to Interrogate		
18260	2026	All other reportable incidents	47265	Dual-chamber implantable pacemaker, rate-responsive			E0402	Hypersensitivity/Allergic Reaction	A24	No Apparent Device Problem	16/03/2026	16/03/2026
							E1714	Rash	A24	No Apparent Device Problem	16/03/2026	16/03/2026
							F08	Hospitalization or Prolonged Hospitalization	A24	No Apparent Device Problem	16/03/2026	16/03/2026
							F1903	Device Explantation	A24	No Apparent Device Problem	16/03/2026	16/03/2026
18261	2026	All other reportable incidents	47265	Dual-chamber implantable pacemaker, rate-responsive			E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
18262	2026	All other reportable incidents	47267	Single-chamber implantable pacemaker, rate-responsive			E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	04/02/2026	04/02/2026
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	04/02/2026	04/02/2026
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	04/02/2026	04/02/2026
18263	2026	All other reportable incidents	47265	Dual-chamber implantable pacemaker, rate-responsive			E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	01/06/2026	01/06/2026
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	01/06/2026	01/06/2026
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	01/06/2026	01/06/2026
18264	2026	All other reportable incidents	47265	Dual-chamber implantable pacemaker, rate-responsive			E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	01/06/2026	01/06/2026
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	01/06/2026	01/06/2026
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	01/06/2026	01/06/2026
18265	2026	All other reportable incidents	47265	Dual-chamber implantable pacemaker, rate-responsive			E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
18266	2026	All other reportable incidents	47265	Dual-chamber implantable pacemaker, rate-responsive			E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	12/01/2026	12/01/2026
							F1901	Additional Surgery	A24	No Apparent Device Problem	12/01/2026	12/01/2026
18267	2026	All other reportable incidents	47265	Dual-chamber implantable pacemaker, rate-responsive			E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	25/03/2026	25/03/2026
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	25/03/2026	25/03/2026
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	25/03/2026	25/03/2026
18268	2026	All other reportable incidents	47267	Single-chamber implantable pacemaker, rate-responsive			E2403	No Clinical Signs, Symptoms or Conditions	A0711	Interrogation Problem		
							F26	No Health Consequences or Impact	A0711	Interrogation Problem		
18269	2026	All other reportable incidents	47265	Dual-chamber implantable pacemaker, rate-responsive			E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
18270	2026	All other reportable incidents	47265	Dual-chamber implantable pacemaker, rate-responsive			E2403	No Clinical Signs, Symptoms or Conditions	A071102	Failure to Interrogate		
									A090811	Unable to Obtain Readings		
							F26	No Health Consequences or Impact	A071102	Failure to Interrogate		
									A090811	Unable to Obtain Readings		
18271	2026	All other reportable incidents	47265	Dual-chamber implantable pacemaker, rate-responsive			E2403	No Clinical Signs, Symptoms or Conditions	A071102	Failure to Interrogate		
									A090811	Unable to Obtain Readings		
							F26	No Health Consequences or Impact	A071102	Failure to Interrogate		
									A090811	Unable to Obtain Readings		
18272	2026	All other reportable incidents	47265	Dual-chamber implantable pacemaker, rate-responsive			E2403	No Clinical Signs, Symptoms or Conditions	A071102	Failure to Interrogate		
							F26	No Health Consequences or Impact	A071102	Failure to Interrogate		
18273	2026	All other reportable incidents	47267	Single-chamber implantable pacemaker, rate-responsive			E2403	No Clinical Signs, Symptoms or Conditions	A071102	Failure to Interrogate		
							F26	No Health Consequences or Impact	A071102	Failure to Interrogate		
18274	2026	All other reportable incidents	47265	Dual-chamber implantable pacemaker, rate-responsive			E2403	No Clinical Signs, Symptoms or Conditions	A1605	Reset Problem		
							F26	No Health Consequences or Impact	A1605	Reset Problem		
18275	2026	All other reportable incidents	47267	Single-chamber implantable pacemaker, rate-responsive			E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem		
							F26	No Health Consequences or Impact	A0705	Battery Problem		
18276	2026		47265	Dual-chamber implantable pacemaker, rate-responsive	CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A0723	Unexpected Mode Switch		
							F26	No Health Consequences or Impact	A0723	Unexpected Mode Switch		
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A0723	Unexpected Mode Switch		
							F26	No Health Consequences or Impact	A0723	Unexpected Mode Switch		
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A0723	Unexpected Mode Switch		
							F26	No Health Consequences or Impact	A0723	Unexpected Mode Switch		
18277	2026	All other reportable incidents	47267	Single-chamber implantable pacemaker, rate-responsive			E2403	No Clinical Signs, Symptoms or Conditions	A072001	Electromagnetic Interference		
							F26	No Health Consequences or Impact	A072001	Electromagnetic Interference		
18278	2026		47265	Dual-chamber implantable pacemaker, rate-responsive	CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A27	Appropriate Device Problem Term/Code not Available		
							F28	Appropriate Health Impact Term/Code not Available	A27	Appropriate Device Problem Term/Code not Available		
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A27	Appropriate Device Problem Term/Code not Available		
							F28	Appropriate Health Impact Term/Code not Available	A27	Appropriate Device Problem Term/Code not Available		
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A27	Appropriate Device Problem Term/Code not Available		

							F28	Appropriate Health Impact Term/Code not Available	A27	Appropriate Device Problem Term/Code not Available		
18279	2026		47265	Dual-chamber implantable pacemaker, rate-responsive	CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A27	Appropriate Device Problem Term/Code not Available		
							F28	Appropriate Health Impact Term/Code not Available	A27	Appropriate Device Problem Term/Code not Available		
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A27	Appropriate Device Problem Term/Code not Available		
							F28	Appropriate Health Impact Term/Code not Available	A27	Appropriate Device Problem Term/Code not Available		
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A27	Appropriate Device Problem Term/Code not Available		
							F28	Appropriate Health Impact Term/Code not Available	A27	Appropriate Device Problem Term/Code not Available		
18280	2026		47265	Dual-chamber implantable pacemaker, rate-responsive	CT1685	Cardiac pacemakers	E0101	Balance Problems	A071205	Pacemaker Found in Back-Up Mode		
									A0909	Unexpected Therapeutic Results		
							E020201	Anxiety	A071205	Pacemaker Found in Back-Up Mode		
									A0909	Unexpected Therapeutic Results		
							F01	Change in Therapeutic Response	A071205	Pacemaker Found in Back-Up Mode		
									A0909	Unexpected Therapeutic Results		
							F0101	Therapeutic Response Decreased	A071205	Pacemaker Found in Back-Up Mode		
									A0909	Unexpected Therapeutic Results		
							F1901	Additional Surgery	A071205	Pacemaker Found in Back-Up Mode		
									A0909	Unexpected Therapeutic Results		
					CT257	Implantable cardiac pacemakers	E0101	Balance Problems	A071205	Pacemaker Found in Back-Up Mode		
									A0909	Unexpected Therapeutic Results		
							E020201	Anxiety	A071205	Pacemaker Found in Back-Up Mode		
									A0909	Unexpected Therapeutic Results		
							F01	Change in Therapeutic Response	A071205	Pacemaker Found in Back-Up Mode		
									A0909	Unexpected Therapeutic Results		
							F0101	Therapeutic Response Decreased	A071205	Pacemaker Found in Back-Up Mode		
									A0909	Unexpected Therapeutic Results		
							F1901	Additional Surgery	A071205	Pacemaker Found in Back-Up Mode		
									A0909	Unexpected Therapeutic Results		
					CT2866	Implantable cardiac electrical stimulators	E0101	Balance Problems	A071205	Pacemaker Found in Back-Up Mode		
									A0909	Unexpected Therapeutic Results		
							E020201	Anxiety	A071205	Pacemaker Found in Back-Up Mode		
									A0909	Unexpected Therapeutic Results		
							F01	Change in Therapeutic Response	A071205	Pacemaker Found in Back-Up Mode		
									A0909	Unexpected Therapeutic Results		
							F0101	Therapeutic Response Decreased	A071205	Pacemaker Found in Back-Up Mode		
									A0909	Unexpected Therapeutic Results		
							F1901	Additional Surgery	A071205	Pacemaker Found in Back-Up Mode		
									A0909	Unexpected Therapeutic Results		
18281	2026	All other reportable incidents	47265	Dual-chamber implantable pacemaker, rate-responsive			E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
18282	2026	All other reportable incidents	47265	Dual-chamber implantable pacemaker, rate-responsive			E0717	Dyspnea	A0908	Incorrect, Inadequate or Imprecise Result or Readings		
							F26	No Health Consequences or Impact	A0908	Incorrect, Inadequate or Imprecise Result or Readings		
18283	2026	All other reportable incidents	47267	Single-chamber implantable pacemaker, rate-responsive			E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
18284	2026	All other reportable incidents	47265	Dual-chamber implantable pacemaker, rate-responsive			E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
18285	2026	All other reportable incidents	47265	Dual-chamber implantable pacemaker, rate-responsive			E2403	No Clinical Signs, Symptoms or Conditions	A071102	Failure to Interrogate		
							F26	No Health Consequences or Impact	A071102	Failure to Interrogate		
18286	2026	All other reportable incidents	47265	Dual-chamber implantable pacemaker, rate-responsive			E2403	No Clinical Signs, Symptoms or Conditions	A071102	Failure to Interrogate		
							F26	No Health Consequences or Impact	A071102	Failure to Interrogate		
18287	2026	All other reportable incidents	47265	Dual-chamber implantable pacemaker, rate-responsive			E2403	No Clinical Signs, Symptoms or Conditions	A160105	False Alarm		
							F26	No Health Consequences or Impact	A160105	False Alarm		
18288	2026	All other reportable incidents	47265	Dual-chamber implantable pacemaker, rate-responsive			E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
18289	2026	All other reportable incidents	47265	Dual-chamber implantable pacemaker, rate-responsive			E2403	No Clinical Signs, Symptoms or Conditions	A071102	Failure to Interrogate		
							F26	No Health Consequences or Impact	A071102	Failure to Interrogate		
18290	2026	Unanticipated serious deterioration in state of health	47265	Dual-chamber implantable pacemaker, rate-responsive			E0402	Hypersensitivity/Allergic Reaction	A24	No Apparent Device Problem	15/06/2023	15/06/2023
							E233001	Chest Pain	A24	No Apparent Device Problem	15/06/2023	15/06/2023
							F08	Hospitalization or Prolonged Hospitalization	A24	No Apparent Device Problem	15/06/2023	15/06/2023
							F12	Serious Injury/ Illness/ Impairment	A24	No Apparent Device Problem	15/06/2023	15/06/2023
							F1903	Device Explantation	A24	No Apparent Device Problem	15/06/2023	15/06/2023
18291	2026	All other reportable incidents	47265	Dual-chamber implantable pacemaker, rate-responsive			E2403	No Clinical Signs, Symptoms or Conditions	A071102	Failure to Interrogate		
							F26	No Health Consequences or Impact	A071102	Failure to Interrogate		
18292	2026	All other reportable incidents	47265	Dual-chamber implantable pacemaker, rate-responsive			E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	22/04/2026	22/04/2026
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	22/04/2026	22/04/2026
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	22/04/2026	22/04/2026
18293	2026	All other reportable incidents	47265	Dual-chamber implantable pacemaker, rate-responsive			E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
18294	2026	All other reportable incidents	47267	Single-chamber implantable pacemaker, rate-responsive			E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
18295	2026	All other reportable incidents	47265	Dual-chamber implantable pacemaker, rate-responsive			E2403	No Clinical Signs, Symptoms or Conditions	A070102	High Capture Threshold		
									A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070102	High Capture Threshold		
									A070504	Premature Discharge of Battery		
18296	2026	All other reportable incidents	47265	Dual-chamber implantable pacemaker, rate-responsive			E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
18297	2026	Unanticipated serious deterioration in state of health	47265	Dual-chamber implantable pacemaker, rate-responsive			E0611	Heart Failure/Congestive Heart Failure	A24	No Apparent Device Problem		
							E0717	Dyspnea	A24	No Apparent Device Problem		
							E0731	Pleural Effusion	A24	No Apparent Device Problem		
							F12	Serious Injury/ Illness/ Impairment	A24	No Apparent Device Problem		
							F2303	Medication Required	A24	No Apparent Device Problem		
							F2307	Reprogramming of Device	A24	No Apparent Device Problem		

18298	2026	All other reportable incidents	47265	Dual-chamber implantable pacemaker, rate-responsive			E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	01/03/2026	01/03/2026
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	01/03/2026	01/03/2026
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	01/03/2026	01/03/2026
18299	2026	All other reportable incidents	47265	Dual-chamber implantable pacemaker, rate-responsive			E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
18300	2026	All other reportable incidents	47265	Dual-chamber implantable pacemaker, rate-responsive			E2312	Fatigue	A0908	Incorrect, Inadequate or Imprecise Result or Readings		
							F08	Hospitalization or Prolonged Hospitalization	A0908	Incorrect, Inadequate or Imprecise Result or Readings		
18301	2026	All other reportable incidents	47265	Dual-chamber implantable pacemaker, rate-responsive			E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	16/01/2026	16/01/2026
							F1901	Additional Surgery	A24	No Apparent Device Problem	16/01/2026	16/01/2026
18302	2026	All other reportable incidents	47265	Dual-chamber implantable pacemaker, rate-responsive			E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
18303	2026	All other reportable incidents	47265	Dual-chamber implantable pacemaker, rate-responsive			E2403	No Clinical Signs, Symptoms or Conditions	A071102	Failure to Interrogate		
							F26	No Health Consequences or Impact	A071102	Failure to Interrogate		
18304	2026	All other reportable incidents	47265	Dual-chamber implantable pacemaker, rate-responsive			E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	09/01/2026	09/01/2026
							F1901	Additional Surgery	A24	No Apparent Device Problem	09/01/2026	09/01/2026
18305	2026	All other reportable incidents	47265	Dual-chamber implantable pacemaker, rate-responsive			E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
18306	2026	All other reportable incidents	47265	Dual-chamber implantable pacemaker, rate-responsive			E2403	No Clinical Signs, Symptoms or Conditions	A0908	Incorrect, Inadequate or Imprecise Result or Readings	25/03/2026	25/03/2026
							F08	Hospitalization or Prolonged Hospitalization	A0908	Incorrect, Inadequate or Imprecise Result or Readings	25/03/2026	25/03/2026
							F1905	Device Revision or Replacement	A0908	Incorrect, Inadequate or Imprecise Result or Readings	25/03/2026	25/03/2026
18307	2026	All other reportable incidents	47265	Dual-chamber implantable pacemaker, rate-responsive			E2403	No Clinical Signs, Symptoms or Conditions	A070102	High Capture Threshold		
									A070909	Oversensing		
									A070910	Undersensing		
									A090801	Signal Artifact/Noise		
									A1301	Failure to Read Input Signal		
							F26	No Health Consequences or Impact	A070102	High Capture Threshold		
									A070909	Oversensing		
									A070910	Undersensing		
									A090801	Signal Artifact/Noise		
									A1301	Failure to Read Input Signal		
18308	2026	All other reportable incidents	47265	Dual-chamber implantable pacemaker, rate-responsive			E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
18309	2026	All other reportable incidents	47265	Dual-chamber implantable pacemaker, rate-responsive			E2403	No Clinical Signs, Symptoms or Conditions	A0908	Incorrect, Inadequate or Imprecise Result or Readings		
							F26	No Health Consequences or Impact	A0908	Incorrect, Inadequate or Imprecise Result or Readings		
18310	2026	All other reportable incidents	47265	Dual-chamber implantable pacemaker, rate-responsive			E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	05/02/2026	05/02/2026
							F1901	Additional Surgery	A24	No Apparent Device Problem	05/02/2026	05/02/2026
18311	2026	All other reportable incidents	47265	Dual-chamber implantable pacemaker, rate-responsive			E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	05/01/2026	05/01/2026
							F1901	Additional Surgery	A24	No Apparent Device Problem	05/01/2026	05/01/2026
18312	2026		47263	Cardiac resynchronization therapy implantable pacemaker	CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A0201	Product Quality Problem		
									A070907	Loss of Threshold		
									A072201	High Impedance		
							F1905	Device Revision or Replacement	A0201	Product Quality Problem		
									A070907	Loss of Threshold		
									A072201	High Impedance		
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A0201	Product Quality Problem		
									A070907	Loss of Threshold		
									A072201	High Impedance		
							F1905	Device Revision or Replacement	A0201	Product Quality Problem		
									A070907	Loss of Threshold		
									A072201	High Impedance		
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A0201	Product Quality Problem		
									A070907	Loss of Threshold		
									A072201	High Impedance		
							F1905	Device Revision or Replacement	A0201	Product Quality Problem		
									A070907	Loss of Threshold		
									A072201	High Impedance		
18313	2026		47267	Single-chamber implantable pacemaker, rate-responsive	CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A070907	Loss of Threshold		
									A072201	High Impedance		
							F1905	Device Revision or Replacement	A070907	Loss of Threshold		
									A072201	High Impedance		
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A070907	Loss of Threshold		
									A072201	High Impedance		
							F1905	Device Revision or Replacement	A070907	Loss of Threshold		
									A072201	High Impedance		
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A070907	Loss of Threshold		
									A072201	High Impedance		
							F1905	Device Revision or Replacement	A070907	Loss of Threshold		
									A072201	High Impedance		
18314	2026	All other reportable incidents	47265	Dual-chamber implantable pacemaker, rate-responsive	CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	27/03/2026	27/03/2026
									A0711	Interrogation Problem	27/03/2026	27/03/2026
							F1901	Additional Surgery	A070504	Premature Discharge of Battery	27/03/2026	27/03/2026
									A0711	Interrogation Problem	27/03/2026	27/03/2026
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	27/03/2026	27/03/2026
									A0711	Interrogation Problem	27/03/2026	27/03/2026
							F1901	Additional Surgery	A070504	Premature Discharge of Battery	27/03/2026	27/03/2026
									A0711	Interrogation Problem	27/03/2026	27/03/2026
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	27/03/2026	27/03/2026
									A0711	Interrogation Problem	27/03/2026	27/03/2026
							F1901	Additional Surgery	A070504	Premature Discharge of Battery	27/03/2026	27/03/2026
									A0711	Interrogation Problem	27/03/2026	27/03/2026
18315	2026	All other reportable incidents	47265	Dual-chamber implantable pacemaker, rate-responsive			E011903	Syncope/Fainting	A070101	Failure to Capture		
							F23	Unexpected Medical Intervention	A070101	Failure to Capture		

18316	2026	All other reportable incidents	47265	Dual-chamber implantable pacemaker, rate-responsive			E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
18317	2026	All other reportable incidents	47263	Cardiac resynchronization therapy implantable pacemaker	CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
18318	2026	All other reportable incidents	47263	Cardiac resynchronization therapy implantable pacemaker	CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	01/05/2026	01/05/2026
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	01/05/2026	01/05/2026
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	01/05/2026	01/05/2026
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	01/05/2026	01/05/2026
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	01/05/2026	01/05/2026
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	01/05/2026	01/05/2026
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	01/05/2026	01/05/2026
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	01/05/2026	01/05/2026
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	01/05/2026	01/05/2026
18319	2026	All other reportable incidents	47265	Dual-chamber implantable pacemaker, rate-responsive			E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
18320	2026	All other reportable incidents	47265	Dual-chamber implantable pacemaker, rate-responsive			E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
18321	2026	All other reportable incidents	47265	Dual-chamber implantable pacemaker, rate-responsive			E2403	No Clinical Signs, Symptoms or Conditions	A0908	Incorrect, Inadequate or Imprecise Result or Readings	26/03/2026	26/03/2026
							F08	Hospitalization or Prolonged Hospitalization	A0908	Incorrect, Inadequate or Imprecise Result or Readings	26/03/2026	26/03/2026
							F1905	Device Revision or Replacement	A0908	Incorrect, Inadequate or Imprecise Result or Readings	26/03/2026	26/03/2026
18322	2026		37265	Dual-chamber implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem		
							F26	No Health Consequences or Impact	A0705	Battery Problem		
18323	2026	All other reportable incidents	47265	Dual-chamber implantable pacemaker, rate-responsive			E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
18324	2026		47265	Dual-chamber implantable pacemaker, rate-responsive	CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A0202	Defective Component		
									A070501	Battery Problem: High Impedance		
							F25	Unanticipated Adverse Device Effect	A0202	Defective Component		
									A070501	Battery Problem: High Impedance		
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A0202	Defective Component		
									A070501	Battery Problem: High Impedance		
							F25	Unanticipated Adverse Device Effect	A0202	Defective Component		
									A070501	Battery Problem: High Impedance		
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A0202	Defective Component		
									A070501	Battery Problem: High Impedance		
							F25	Unanticipated Adverse Device Effect	A0202	Defective Component		
									A070501	Battery Problem: High Impedance		
18325	2026	All other reportable incidents	47265	Dual-chamber implantable pacemaker, rate-responsive			E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
18326	2026	All other reportable incidents	47265	Dual-chamber implantable pacemaker, rate-responsive	CT1685	Cardiac pacemakers	E020201	Anxiety	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT257	Implantable cardiac pacemakers	E020201	Anxiety	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT2866	Implantable cardiac electrical stimulators	E020201	Anxiety	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
18327	2026	All other reportable incidents	47265	Dual-chamber implantable pacemaker, rate-responsive			E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
18328	2026	All other reportable incidents	47265	Dual-chamber implantable pacemaker, rate-responsive			E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
18329	2026	All other reportable incidents	47265	Dual-chamber implantable pacemaker, rate-responsive			E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
18330	2026	All other reportable incidents	47265	Dual-chamber implantable pacemaker, rate-responsive	CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
18331	2026	All other reportable incidents	47265	Dual-chamber implantable pacemaker, rate-responsive	CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
18332	2026	All other reportable incidents	47265	Dual-chamber implantable pacemaker, rate-responsive	CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
18333	2026		47270	Cardiac resynchronization therapy implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A160602	Premature End-of-Life Indicator	26/02/2026	
							F08	Hospitalization or Prolonged Hospitalization	A160602	Premature End-of-Life Indicator	26/02/2026	
							F26	No Health Consequences or Impact	A160602	Premature End-of-Life Indicator	26/02/2026	
18334	2026	All other reportable incidents	35852	Single-chamber implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E020201	Anxiety	A070504	Premature Discharge of Battery	26/02/2025	26/02/2025
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	26/02/2025	26/02/2025
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	26/02/2025	26/02/2025
18335	2026	All other reportable incidents	47265	Dual-chamber implantable pacemaker, rate-responsive			E011903	Syncope/Fainting	A070102	High Capture Threshold		
									A0709	Device Sensing Problem		
									A0713	Defibrillation/Stimulation Problem		
									A1107	Data Problem		
							F15	Recognized Device or Procedural Complication	A070102	High Capture Threshold		
									A0709	Device Sensing Problem		
									A0713	Defibrillation/Stimulation Problem		

									A1107	Data Problem		
18336	2026	All other reportable incidents	47265	Dual-chamber implantable pacemaker, rate-responsive	CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
18337	2026	All other reportable incidents	47265	Dual-chamber implantable pacemaker, rate-responsive			E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
18338	2026	All other reportable incidents	47265	Dual-chamber implantable pacemaker, rate-responsive	CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
18339	2026	All other reportable incidents	60789	Intracardiac pacemaker	CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A13	Communication or Transmission Problem		
							F26	No Health Consequences or Impact	A13	Communication or Transmission Problem		
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A13	Communication or Transmission Problem		
							F26	No Health Consequences or Impact	A13	Communication or Transmission Problem		
					CT2818	Catheterization/Cannulation/Fluid path devices	E2403	No Clinical Signs, Symptoms or Conditions	A13	Communication or Transmission Problem		
							F26	No Health Consequences or Impact	A13	Communication or Transmission Problem		
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A13	Communication or Transmission Problem		
							F26	No Health Consequences or Impact	A13	Communication or Transmission Problem		
18340	2026		37265	Dual-chamber implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
									A09	Output Problem		
									A160601	Premature Elective Replacement Indicator		
							F23	Unexpected Medical Intervention	A070504	Premature Discharge of Battery		
									A09	Output Problem		
									A160601	Premature Elective Replacement Indicator		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
									A09	Output Problem		
									A160601	Premature Elective Replacement Indicator		
18341	2026		47265	Dual-chamber implantable pacemaker, rate-responsive			E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	24/03/2026	24/03/2026
							F1901	Additional Surgery	A24	No Apparent Device Problem	24/03/2026	24/03/2026
18342	2026	All other reportable incidents	47265	Dual-chamber implantable pacemaker, rate-responsive	CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
18343	2026	All other reportable incidents	47263	Cardiac resynchronization therapy implantable pacemaker	CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
18344	2026	All other reportable incidents	47265	Dual-chamber implantable pacemaker, rate-responsive	CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A0908	Incorrect, Inadequate or Imprecise Result or Readings	09/04/2026	09/04/2026
							F08	Hospitalization or Prolonged Hospitalization	A0908	Incorrect, Inadequate or Imprecise Result or Readings	09/04/2026	09/04/2026
							F1905	Device Revision or Replacement	A0908	Incorrect, Inadequate or Imprecise Result or Readings	09/04/2026	09/04/2026
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A0908	Incorrect, Inadequate or Imprecise Result or Readings	09/04/2026	09/04/2026
							F08	Hospitalization or Prolonged Hospitalization	A0908	Incorrect, Inadequate or Imprecise Result or Readings	09/04/2026	09/04/2026
							F1905	Device Revision or Replacement	A0908	Incorrect, Inadequate or Imprecise Result or Readings	09/04/2026	09/04/2026
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A0908	Incorrect, Inadequate or Imprecise Result or Readings	09/04/2026	09/04/2026
							F08	Hospitalization or Prolonged Hospitalization	A0908	Incorrect, Inadequate or Imprecise Result or Readings	09/04/2026	09/04/2026
							F1905	Device Revision or Replacement	A0908	Incorrect, Inadequate or Imprecise Result or Readings	09/04/2026	09/04/2026
18345	2026	All other reportable incidents	47265	Dual-chamber implantable pacemaker, rate-responsive	CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
18346	2026	All other reportable incidents	47265	Dual-chamber implantable pacemaker, rate-responsive			E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
18347	2026	All other reportable incidents	47265	Dual-chamber implantable pacemaker, rate-responsive	CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
18348	2026	All other reportable incidents	47265	Dual-chamber implantable pacemaker, rate-responsive	CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		

18349	2026	All other reportable incidents	47265	Dual-chamber implantable pacemaker, rate-responsive	CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
18350	2026	All other reportable incidents	47265	Dual-chamber implantable pacemaker, rate-responsive	CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
18351	2026		35224	External pacemaker, epicardial pacing	CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A040502	Corroded		
									A0705	Battery Problem		
							F26	No Health Consequences or Impact	A040502	Corroded		
									A0705	Battery Problem		
18352	2026	All other reportable incidents	47265	Dual-chamber implantable pacemaker, rate-responsive	CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	14/04/2026	14/04/2026
							F1901	Additional Surgery	A070504	Premature Discharge of Battery	14/04/2026	14/04/2026
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	14/04/2026	14/04/2026
							F1901	Additional Surgery	A070504	Premature Discharge of Battery	14/04/2026	14/04/2026
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	14/04/2026	14/04/2026
							F1901	Additional Surgery	A070504	Premature Discharge of Battery	14/04/2026	14/04/2026
18353	2026	All other reportable incidents	47265	Dual-chamber implantable pacemaker, rate-responsive			E020201	Anxiety	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
18354	2026	All other reportable incidents	47265	Dual-chamber implantable pacemaker, rate-responsive			E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
18355	2026	All other reportable incidents	47265	Dual-chamber implantable pacemaker, rate-responsive	CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	06/07/2026	06/07/2026
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	06/07/2026	06/07/2026
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	06/07/2026	06/07/2026
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	06/07/2026	06/07/2026
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	06/07/2026	06/07/2026
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	06/07/2026	06/07/2026
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	06/07/2026	06/07/2026
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	06/07/2026	06/07/2026
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	06/07/2026	06/07/2026
18356	2026	All other reportable incidents	47267	Single-chamber implantable pacemaker, rate-responsive			E2403	No Clinical Signs, Symptoms or Conditions	A0711	Interrogation Problem		
							F26	No Health Consequences or Impact	A0711	Interrogation Problem		
18357	2026		17882	Physiologic-monitoring defibrillation system	CT1684	Cardiac pacemakers and associated devices	E2401	Insufficient Information	A0201	Product Quality Problem		
									A070912	Incorrect Interpretation of Signal		
									A090101	Inappropriate Audible Prompt/Feedback		
							F05	Delay to Treatment/ Therapy	A0201	Product Quality Problem		
									A070912	Incorrect Interpretation of Signal		
									A090101	Inappropriate Audible Prompt/Feedback		
18358	2026	All other reportable incidents	47265	Dual-chamber implantable pacemaker, rate-responsive			E0717	Dyspnea	A0712	Pacing Problem		
							F11	Non-Serious Injury/ Illness/ Impairment	A0712	Pacing Problem		
18359	2026	All other reportable incidents	47265	Dual-chamber implantable pacemaker, rate-responsive			E2403	No Clinical Signs, Symptoms or Conditions	A0712	Pacing Problem		
							F26	No Health Consequences or Impact	A0712	Pacing Problem		
18360	2026		47265	Dual-chamber implantable pacemaker, rate-responsive	CT1685	Cardiac pacemakers	E020201	Anxiety	A0201	Product Quality Problem		
									A2205	Device Provides Uninformative Message		
							F1901	Additional Surgery	A0201	Product Quality Problem		
									A2205	Device Provides Uninformative Message		
					CT257	Implantable cardiac pacemakers	E020201	Anxiety	A0201	Product Quality Problem		
									A2205	Device Provides Uninformative Message		
							F1901	Additional Surgery	A0201	Product Quality Problem		
									A2205	Device Provides Uninformative Message		
					CT2866	Implantable cardiac electrical stimulators	E020201	Anxiety	A0201	Product Quality Problem		
									A2205	Device Provides Uninformative Message		
							F1901	Additional Surgery	A0201	Product Quality Problem		
									A2205	Device Provides Uninformative Message		
18361	2026	All other reportable incidents	47265	Dual-chamber implantable pacemaker, rate-responsive			E2403	No Clinical Signs, Symptoms or Conditions	A070101	Failure to Capture		
									A070909	Oversensing		
									A090801	Signal Artifact/Noise		
							F2203	Imaging Required	A070101	Failure to Capture		
									A070909	Oversensing		
									A090801	Signal Artifact/Noise		
							F26	No Health Consequences or Impact	A070101	Failure to Capture		
									A070909	Oversensing		
									A090801	Signal Artifact/Noise		
18362	2026	All other reportable incidents	47265	Dual-chamber implantable pacemaker, rate-responsive	CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
18363	2026	All other reportable incidents	47265	Dual-chamber implantable pacemaker, rate-responsive			E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
18364	2026	All other reportable incidents	47265	Dual-chamber implantable pacemaker, rate-responsive			E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
18365	2026	All other reportable incidents	47267	Single-chamber implantable pacemaker, rate-responsive	CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
18366	2026	All other reportable incidents	47265	Dual-chamber implantable pacemaker, rate-responsive			E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
18367	2026	All other reportable incidents	47265	Dual-chamber implantable pacemaker, rate-responsive	CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		

					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
18368	2026	All other reportable incidents	47265	Dual-chamber implantable pacemaker, rate-responsive	CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
18369	2026		47265	Dual-chamber implantable pacemaker, rate-responsive	CT1685	Cardiac pacemakers	E020201	Anxiety	A0201	Product Quality Problem		
									A071205	Pacemaker Found in Back-Up Mode		
							F1901	Additional Surgery	A0201	Product Quality Problem		
									A071205	Pacemaker Found in Back-Up Mode		
					CT257	Implantable cardiac pacemakers	E020201	Anxiety	A0201	Product Quality Problem		
									A071205	Pacemaker Found in Back-Up Mode		
							F1901	Additional Surgery	A0201	Product Quality Problem		
									A071205	Pacemaker Found in Back-Up Mode		
					CT2866	Implantable cardiac electrical stimulators	E020201	Anxiety	A0201	Product Quality Problem		
									A071205	Pacemaker Found in Back-Up Mode		
							F1901	Additional Surgery	A0201	Product Quality Problem		
									A071205	Pacemaker Found in Back-Up Mode		
18370	2026	All other reportable incidents	60789	Intracardiac pacemaker			E2403	No Clinical Signs, Symptoms or Conditions	A0701	Capturing Problem		
									A070101	Failure to Capture		
							F12	Serious Injury/ Illness/ Impairment	A0701	Capturing Problem		
									A070101	Failure to Capture		
							F2307	Reprogramming of Device	A0701	Capturing Problem		
									A070101	Failure to Capture		
18371	2026	All other reportable incidents	47265	Dual-chamber implantable pacemaker, rate-responsive			E012207	Twitching	A0908	Incorrect, Inadequate or Imprecise Result or Readings	29/04/2026	29/04/2026
							F08	Hospitalization or Prolonged Hospitalization	A0908	Incorrect, Inadequate or Imprecise Result or Readings	29/04/2026	29/04/2026
							F1905	Device Revision or Replacement	A0908	Incorrect, Inadequate or Imprecise Result or Readings	29/04/2026	29/04/2026
18372	2026	All other reportable incidents	47265	Dual-chamber implantable pacemaker, rate-responsive	CT1685	Cardiac pacemakers	E020201	Anxiety	A070504	Premature Discharge of Battery		
									A0711	Interrogation Problem		
									A2205	Device Provides Uninformative Message		
									A23	Use of Device Problem		
									A27	Appropriate Device Problem Term/Code not Available		
							E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
									A0711	Interrogation Problem		
									A2205	Device Provides Uninformative Message		
									A23	Use of Device Problem		
									A27	Appropriate Device Problem Term/Code not Available		
							F1901	Additional Surgery	A070504	Premature Discharge of Battery		
									A0711	Interrogation Problem		
									A2205	Device Provides Uninformative Message		
									A23	Use of Device Problem		
									A27	Appropriate Device Problem Term/Code not Available		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
									A0711	Interrogation Problem		
									A2205	Device Provides Uninformative Message		
									A23	Use of Device Problem		
									A27	Appropriate Device Problem Term/Code not Available		
					CT257	Implantable cardiac pacemakers	E020201	Anxiety	A070504	Premature Discharge of Battery		
									A0711	Interrogation Problem		
									A2205	Device Provides Uninformative Message		
									A23	Use of Device Problem		
									A27	Appropriate Device Problem Term/Code not Available		
							E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
									A0711	Interrogation Problem		
									A2205	Device Provides Uninformative Message		
									A23	Use of Device Problem		
									A27	Appropriate Device Problem Term/Code not Available		
							F1901	Additional Surgery	A070504	Premature Discharge of Battery		
									A0711	Interrogation Problem		
									A2205	Device Provides Uninformative Message		
									A23	Use of Device Problem		
									A27	Appropriate Device Problem Term/Code not Available		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
									A0711	Interrogation Problem		
									A2205	Device Provides Uninformative Message		
									A23	Use of Device Problem		
									A27	Appropriate Device Problem Term/Code not Available		
					CT2866	Implantable cardiac electrical stimulators	E020201	Anxiety	A070504	Premature Discharge of Battery		
									A0711	Interrogation Problem		
									A2205	Device Provides Uninformative Message		
									A23	Use of Device Problem		
									A27	Appropriate Device Problem Term/Code not Available		
							E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
									A0711	Interrogation Problem		
									A2205	Device Provides Uninformative Message		

									A23	Use of Device Problem		
									A27	Appropriate Device Problem Term/Code not Available		
							F1901	Additional Surgery	A070504	Premature Discharge of Battery		
									A0711	Interrogation Problem		
									A2205	Device Provides Uninformative Message		
									A23	Use of Device Problem		
									A27	Appropriate Device Problem Term/Code not Available		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
									A0711	Interrogation Problem		
									A2205	Device Provides Uninformative Message		
									A23	Use of Device Problem		
									A27	Appropriate Device Problem Term/Code not Available		
18373	2026	All other reportable incidents	47265	Dual-chamber implantable pacemaker, rate-responsive			E0119	Loss of Consciousness	A1702	Device-Device Incompatibility	30/03/2026	30/03/2026
							F1905	Device Revision or Replacement	A1702	Device-Device Incompatibility	30/03/2026	30/03/2026
18374	2026	All other reportable incidents	47265	Dual-chamber implantable pacemaker, rate-responsive	CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
18375	2026	All other reportable incidents	47265	Dual-chamber implantable pacemaker, rate-responsive			E2403	No Clinical Signs, Symptoms or Conditions	A070101	Failure to Capture		
							F1010	Inadequate Treatment	A070101	Failure to Capture		
18376	2026	All other reportable incidents	47265	Dual-chamber implantable pacemaker, rate-responsive	CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
18377	2026	All other reportable incidents	47265	Dual-chamber implantable pacemaker, rate-responsive	CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
18378	2026	All other reportable incidents	47265	Dual-chamber implantable pacemaker, rate-responsive			E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
									A0908	Incorrect, Inadequate or Imprecise Result or Readings		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
									A0908	Incorrect, Inadequate or Imprecise Result or Readings		
18379	2026	All other reportable incidents	37265	Dual-chamber implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E060109	Tachycardia	A0709	Device Sensing Problem		
							F12	Serious Injury/ Illness/ Impairment	A0709	Device Sensing Problem		
							F2307	Reprogramming of Device	A0709	Device Sensing Problem		
18380	2026	All other reportable incidents	47265	Dual-chamber implantable pacemaker, rate-responsive			E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem		
									A0711	Interrogation Problem		
									A0908	Incorrect, Inadequate or Imprecise Result or Readings		
									A1107	Data Problem		
									A23	Use of Device Problem		
							F26	No Health Consequences or Impact	A0705	Battery Problem		
									A0711	Interrogation Problem		
									A0908	Incorrect, Inadequate or Imprecise Result or Readings		
									A1107	Data Problem		
									A23	Use of Device Problem		
18381	2026	All other reportable incidents	47265	Dual-chamber implantable pacemaker, rate-responsive			E2311	Discomfort	A0712	Pacing Problem		
							E2312	Fatigue	A0712	Pacing Problem		
							E233001	Chest Pain	A0712	Pacing Problem		
							F1010	Inadequate Treatment	A0712	Pacing Problem		
18382	2026	All other reportable incidents	47265	Dual-chamber implantable pacemaker, rate-responsive			E2403	No Clinical Signs, Symptoms or Conditions	A070101	Failure to Capture		
							F1010	Inadequate Treatment	A070101	Failure to Capture		
18383	2026	All other reportable incidents	47265	Dual-chamber implantable pacemaker, rate-responsive	CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
18384	2026	All other reportable incidents	47265	Dual-chamber implantable pacemaker, rate-responsive			E012207	Twitching	A0908	Incorrect, Inadequate or Imprecise Result or Readings	08/04/2026	08/04/2026
							E1605	Cramp(s) /Muscle Spasm(s)	A0908	Incorrect, Inadequate or Imprecise Result or Readings	08/04/2026	08/04/2026
							F08	Hospitalization or Prolonged Hospitalization	A0908	Incorrect, Inadequate or Imprecise Result or Readings	08/04/2026	08/04/2026
							F1905	Device Revision or Replacement	A0908	Incorrect, Inadequate or Imprecise Result or Readings	08/04/2026	08/04/2026
18385	2026	All other reportable incidents	47265	Dual-chamber implantable pacemaker, rate-responsive	CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
18386	2026	All other reportable incidents	47265	Dual-chamber implantable pacemaker, rate-responsive	CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		

18387	2026	All other reportable incidents	47265	Dual-chamber implantable pacemaker, rate-responsive	CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
18388	2026	All other reportable incidents	47270	Cardiac resynchronization therapy implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A05	Mechanical Problem		
									A0722	Impedance Problem		
									A072201	High Impedance		
							F08	Hospitalization or Prolonged Hospitalization	A05	Mechanical Problem		
									A0722	Impedance Problem		
									A072201	High Impedance		
							F12	Serious Injury/ Illness/ Impairment	A05	Mechanical Problem		
									A0722	Impedance Problem		
									A072201	High Impedance		
							F1905	Device Revision or Replacement	A05	Mechanical Problem		
									A0722	Impedance Problem		
									A072201	High Impedance		
18389	2026	All other reportable incidents	35852	Single-chamber implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E1901	Bacterial Infection	A24	No Apparent Device Problem	11/06/2026	11/06/2026
							E1906	Unspecified Infection	A24	No Apparent Device Problem	11/06/2026	11/06/2026
							F08	Hospitalization or Prolonged Hospitalization	A24	No Apparent Device Problem	11/06/2026	11/06/2026
							F12	Serious Injury/ Illness/ Impairment	A24	No Apparent Device Problem	11/06/2026	11/06/2026
							F1903	Device Explantation	A24	No Apparent Device Problem	11/06/2026	11/06/2026
							F2303	Medication Required	A24	No Apparent Device Problem	11/06/2026	11/06/2026
18390	2026	All other reportable incidents	47265	Dual-chamber implantable pacemaker, rate-responsive	CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions				
							F26	No Health Consequences or Impact				
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions				
							F26	No Health Consequences or Impact				
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions				
							F26	No Health Consequences or Impact				
18391	2026	All other reportable incidents	47265	Dual-chamber implantable pacemaker, rate-responsive	CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
18392	2026	All other reportable incidents	47265	Dual-chamber implantable pacemaker, rate-responsive	CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
18393	2026	All other reportable incidents	47265	Dual-chamber implantable pacemaker, rate-responsive	CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
18394	2026	All other reportable incidents	47265	Dual-chamber implantable pacemaker, rate-responsive			E2403	No Clinical Signs, Symptoms or Conditions	A1605	Reset Problem		
							F26	No Health Consequences or Impact	A1605	Reset Problem		
18395	2026	All other reportable incidents	47265	Dual-chamber implantable pacemaker, rate-responsive	CT1147	Defibrillators and associated devices	E2401	Insufficient Information	A0203	Defective Device		
									A0908	Incorrect, Inadequate or Imprecise Result or Readings		
									A191002	Loss of Power		
							E2403	No Clinical Signs, Symptoms or Conditions	A0203	Defective Device		
									A0908	Incorrect, Inadequate or Imprecise Result or Readings		
									A191002	Loss of Power		
							F26	No Health Consequences or Impact	A0203	Defective Device		
									A0908	Incorrect, Inadequate or Imprecise Result or Readings		
									A191002	Loss of Power		
							F28	Appropriate Health Impact Term/Code not Available	A0203	Defective Device		
									A0908	Incorrect, Inadequate or Imprecise Result or Readings		
									A191002	Loss of Power		
					CT145	Cardiovascular devices	E2401	Insufficient Information	A0203	Defective Device		
									A0908	Incorrect, Inadequate or Imprecise Result or Readings		
									A191002	Loss of Power		
							E2403	No Clinical Signs, Symptoms or Conditions	A0203	Defective Device		
									A0908	Incorrect, Inadequate or Imprecise Result or Readings		
									A191002	Loss of Power		
							F26	No Health Consequences or Impact	A0203	Defective Device		
									A0908	Incorrect, Inadequate or Imprecise Result or Readings		
									A191002	Loss of Power		
							F28	Appropriate Health Impact Term/Code not Available	A0203	Defective Device		
									A0908	Incorrect, Inadequate or Imprecise Result or Readings		
									A191002	Loss of Power		
					CT1684	Cardiac pacemakers and associated devices	E2401	Insufficient Information	A0203	Defective Device		
									A0908	Incorrect, Inadequate or Imprecise Result or Readings		
									A191002	Loss of Power		
							E2403	No Clinical Signs, Symptoms or Conditions	A0203	Defective Device		
									A0908	Incorrect, Inadequate or Imprecise Result or Readings		
									A191002	Loss of Power		

							F26	No Health Consequences or Impact	A0203	Defective Device		
									A0908	Incorrect, Inadequate or Imprecise Result or Readings		
									A191002	Loss of Power		
							F28	Appropriate Health Impact Term/Code not Available	A0203	Defective Device		
									A0908	Incorrect, Inadequate or Imprecise Result or Readings		
									A191002	Loss of Power		
					CT2727	Active implant devices/ancillaries	E2401	Insufficient Information	A0203	Defective Device		
									A0908	Incorrect, Inadequate or Imprecise Result or Readings		
									A191002	Loss of Power		
							E2403	No Clinical Signs, Symptoms or Conditions	A0203	Defective Device		
									A0908	Incorrect, Inadequate or Imprecise Result or Readings		
									A191002	Loss of Power		
							F26	No Health Consequences or Impact	A0203	Defective Device		
									A0908	Incorrect, Inadequate or Imprecise Result or Readings		
									A191002	Loss of Power		
							F28	Appropriate Health Impact Term/Code not Available	A0203	Defective Device		
									A0908	Incorrect, Inadequate or Imprecise Result or Readings		
									A191002	Loss of Power		
					CT2750	Implantable cardiac electrical stimulators and associated devices	E2401	Insufficient Information	A0203	Defective Device		
									A0908	Incorrect, Inadequate or Imprecise Result or Readings		
									A191002	Loss of Power		
							E2403	No Clinical Signs, Symptoms or Conditions	A0203	Defective Device		
									A0908	Incorrect, Inadequate or Imprecise Result or Readings		
									A191002	Loss of Power		
							F26	No Health Consequences or Impact	A0203	Defective Device		
									A0908	Incorrect, Inadequate or Imprecise Result or Readings		
									A191002	Loss of Power		
							F28	Appropriate Health Impact Term/Code not Available	A0203	Defective Device		
									A0908	Incorrect, Inadequate or Imprecise Result or Readings		
									A191002	Loss of Power		
18396	2026	Death	47265	Dual-chamber implantable pacemaker, rate-responsive			E060104	Bradycardia	A0712	Pacing Problem	01/06/2017	30/06/2017
							E0602	Cardiac Arrest	A0712	Pacing Problem	01/06/2017	30/06/2017
							E0717	Dyspnea	A0712	Pacing Problem	01/06/2017	30/06/2017
							E2321	Low Blood Pressure/ Hypotension	A0712	Pacing Problem	01/06/2017	30/06/2017
							F02	Death	A0712	Pacing Problem	01/06/2017	30/06/2017
18397	2026	All other reportable incidents	47265	Dual-chamber implantable pacemaker, rate-responsive	CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
18398	2026	All other reportable incidents	47265	Dual-chamber implantable pacemaker, rate-responsive	CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	26/06/2026	26/06/2026
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	26/06/2026	26/06/2026
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	26/06/2026	26/06/2026
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	26/06/2026	26/06/2026
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	26/06/2026	26/06/2026
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	26/06/2026	26/06/2026
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	26/06/2026	26/06/2026
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	26/06/2026	26/06/2026
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	26/06/2026	26/06/2026
18399	2026	All other reportable incidents	47265	Dual-chamber implantable pacemaker, rate-responsive			E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	08/04/2026	08/04/2026
									A0711	Interrogation Problem	08/04/2026	08/04/2026
									A0712	Pacing Problem	08/04/2026	08/04/2026
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	08/04/2026	08/04/2026
									A0711	Interrogation Problem	08/04/2026	08/04/2026
									A0712	Pacing Problem	08/04/2026	08/04/2026
							F1001	Absence of Treatment	A070504	Premature Discharge of Battery	08/04/2026	08/04/2026
									A0711	Interrogation Problem	08/04/2026	08/04/2026
									A0712	Pacing Problem	08/04/2026	08/04/2026
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	08/04/2026	08/04/2026
									A0711	Interrogation Problem	08/04/2026	08/04/2026
									A0712	Pacing Problem	08/04/2026	08/04/2026
18400	2026	All other reportable incidents	47265	Dual-chamber implantable pacemaker, rate-responsive	CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
18401	2026	All other reportable incidents	47265	Dual-chamber implantable pacemaker, rate-responsive	CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
18402	2026	All other reportable incidents	47265	Dual-chamber implantable pacemaker, rate-responsive			E2403	No Clinical Signs, Symptoms or Conditions	A0712	Pacing Problem		
							F26	No Health Consequences or Impact	A0712	Pacing Problem		
18403	2026	All other reportable incidents	47265	Dual-chamber implantable pacemaker, rate-responsive	CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		

					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
18404	2026	All other reportable incidents	47265	Dual-chamber implantable pacemaker, rate-responsive	CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
18405	2026	All other reportable incidents	47265	Dual-chamber implantable pacemaker, rate-responsive	CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	02/07/2026	02/07/2026
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	02/07/2026	02/07/2026
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	02/07/2026	02/07/2026
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	02/07/2026	02/07/2026
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	02/07/2026	02/07/2026
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	02/07/2026	02/07/2026
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	02/07/2026	02/07/2026
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	02/07/2026	02/07/2026
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	02/07/2026	02/07/2026
18406	2026	All other reportable incidents	47265	Dual-chamber implantable pacemaker, rate-responsive	CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
18407	2026	All other reportable incidents	37265	Dual-chamber implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A070909	Oversensing		
							F26	No Health Consequences or Impact	A070909	Oversensing		
18408	2026		47267	Single-chamber implantable pacemaker, rate-responsive	CT1685	Cardiac pacemakers	E2401	Insufficient Information	A24	No Apparent Device Problem		
									A27	Appropriate Device Problem Term/Code not Available		
							F02	Death	A24	No Apparent Device Problem		
									A27	Appropriate Device Problem Term/Code not Available		
					CT257	Implantable cardiac pacemakers	E2401	Insufficient Information	A24	No Apparent Device Problem		
									A27	Appropriate Device Problem Term/Code not Available		
							F02	Death	A24	No Apparent Device Problem		
									A27	Appropriate Device Problem Term/Code not Available		
					CT2866	Implantable cardiac electrical stimulators	E2401	Insufficient Information	A24	No Apparent Device Problem		
									A27	Appropriate Device Problem Term/Code not Available		
							F02	Death	A24	No Apparent Device Problem		
									A27	Appropriate Device Problem Term/Code not Available		
18409	2026	All other reportable incidents	47265	Dual-chamber implantable pacemaker, rate-responsive			E2403	No Clinical Signs, Symptoms or Conditions	A24	No Apparent Device Problem	19/03/2026	19/03/2026
							F1901	Additional Surgery	A24	No Apparent Device Problem	19/03/2026	19/03/2026
18410	2026		47265	Dual-chamber implantable pacemaker, rate-responsive	CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A070501	Battery Problem: High Impedance		
									A071207	Pacing Inadequately		
							F1901	Additional Surgery	A070501	Battery Problem: High Impedance		
									A071207	Pacing Inadequately		
							F26	No Health Consequences or Impact	A070501	Battery Problem: High Impedance		
									A071207	Pacing Inadequately		
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A070501	Battery Problem: High Impedance		
									A071207	Pacing Inadequately		
							F1901	Additional Surgery	A070501	Battery Problem: High Impedance		
									A071207	Pacing Inadequately		
							F26	No Health Consequences or Impact	A070501	Battery Problem: High Impedance		
									A071207	Pacing Inadequately		
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A070501	Battery Problem: High Impedance		
									A071207	Pacing Inadequately		
							F1901	Additional Surgery	A070501	Battery Problem: High Impedance		
									A071207	Pacing Inadequately		
							F26	No Health Consequences or Impact	A070501	Battery Problem: High Impedance		
									A071207	Pacing Inadequately		
18411	2026	All other reportable incidents	47267	Single-chamber implantable pacemaker, rate-responsive	CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions				
							F26	No Health Consequences or Impact				
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions				
							F26	No Health Consequences or Impact				
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions				
							F26	No Health Consequences or Impact				
18412	2026		47270	Cardiac resynchronization therapy implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A0201	Product Quality Problem		
									A070504	Premature Discharge of Battery		
							F08	Hospitalization or Prolonged Hospitalization	A0201	Product Quality Problem		
									A070504	Premature Discharge of Battery		
							F1901	Additional Surgery	A0201	Product Quality Problem		
									A070504	Premature Discharge of Battery		
							F1905	Device Revision or Replacement	A0201	Product Quality Problem		
									A070504	Premature Discharge of Battery		
18413	2026	All other reportable incidents	47265	Dual-chamber implantable pacemaker, rate-responsive			E2403	No Clinical Signs, Symptoms or Conditions	A05	Mechanical Problem	15/04/2026	15/04/2026
									A070909	Oversensing	15/04/2026	15/04/2026
							F26	No Health Consequences or Impact	A05	Mechanical Problem	15/04/2026	15/04/2026
									A070909	Oversensing	15/04/2026	15/04/2026
18414	2026	All other reportable incidents	47265	Dual-chamber implantable pacemaker, rate-responsive	CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
18415	2026	All other reportable incidents	47265	Dual-chamber implantable pacemaker, rate-			E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	09/04/2026	09/04/2026

				responsive					A0908	Incorrect, Inadequate or Imprecise Result or Readings	09/04/2026	09/04/2026
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	09/04/2026	09/04/2026
									A0908	Incorrect, Inadequate or Imprecise Result or Readings	09/04/2026	09/04/2026
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	09/04/2026	09/04/2026
									A0908	Incorrect, Inadequate or Imprecise Result or Readings	09/04/2026	09/04/2026
18416	2026	All other reportable incidents	47265	Dual-chamber implantable pacemaker, rate-responsive	CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
18417	2026	All other reportable incidents	47265	Dual-chamber implantable pacemaker, rate-responsive			E2403	No Clinical Signs, Symptoms or Conditions	A070909	Oversensing		
									A0712	Pacing Problem		
							F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A070909	Oversensing		
									A0712	Pacing Problem		
18418	2026	All other reportable incidents	47265	Dual-chamber implantable pacemaker, rate-responsive			E2403	No Clinical Signs, Symptoms or Conditions	A0711	Interrogation Problem		
									A1107	Data Problem		
									A23	Use of Device Problem		
							F26	No Health Consequences or Impact	A0711	Interrogation Problem		
									A1107	Data Problem		
									A23	Use of Device Problem		
18419	2026	All other reportable incidents	47265	Dual-chamber implantable pacemaker, rate-responsive			E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	17/02/2026	17/02/2026
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	17/02/2026	17/02/2026
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	17/02/2026	17/02/2026
18420	2026	All other reportable incidents	47265	Dual-chamber implantable pacemaker, rate-responsive	CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
18421	2026	All other reportable incidents	17882	Physiologic-monitoring defibrillation system	CT1684	Cardiac pacemakers and associated devices	E2401	Insufficient Information	A071301	Failure to Deliver Shock/Stimulation		
							F24	Insufficient Health Impact Information	A071301	Failure to Deliver Shock/Stimulation		
18422	2026	All other reportable incidents	47265	Dual-chamber implantable pacemaker, rate-responsive	CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions				
							F26	No Health Consequences or Impact				
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions				
							F26	No Health Consequences or Impact				
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions				
							F26	No Health Consequences or Impact				
18423	2026	All other reportable incidents	60789	Intracardiac pacemaker			E2403	No Clinical Signs, Symptoms or Conditions	A070102	High Capture Threshold		
									A070504	Premature Discharge of Battery		
							F12	Serious Injury/ Illness/ Impairment	A070102	High Capture Threshold		
									A070504	Premature Discharge of Battery		
							F2307	Reprogramming of Device	A070102	High Capture Threshold		
									A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070102	High Capture Threshold		
									A070504	Premature Discharge of Battery		
18424	2026	All other reportable incidents	47265	Dual-chamber implantable pacemaker, rate-responsive			E2403	No Clinical Signs, Symptoms or Conditions	A1605	Reset Problem		
							F2307	Reprogramming of Device	A1605	Reset Problem		
18425	2026	All other reportable incidents	47265	Dual-chamber implantable pacemaker, rate-responsive			E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
18426	2026	All other reportable incidents	47265	Dual-chamber implantable pacemaker, rate-responsive			E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
18427	2026	All other reportable incidents	47265	Dual-chamber implantable pacemaker, rate-responsive	CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions				
							F26	No Health Consequences or Impact				
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions				
							F26	No Health Consequences or Impact				
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions				
							F26	No Health Consequences or Impact				
18428	2026	All other reportable incidents	47265	Dual-chamber implantable pacemaker, rate-responsive			E2403	No Clinical Signs, Symptoms or Conditions	A070909	Oversensing		
									A0712	Pacing Problem		
							F26	No Health Consequences or Impact	A070909	Oversensing		
									A0712	Pacing Problem		
18429	2026	All other reportable incidents	47265	Dual-chamber implantable pacemaker, rate-responsive			E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
18430	2026	All other reportable incidents	47265	Dual-chamber implantable pacemaker, rate-responsive	CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions				
							F26	No Health Consequences or Impact				
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions				
							F26	No Health Consequences or Impact				
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions				
							F26	No Health Consequences or Impact				
18431	2026	All other reportable incidents	47265	Dual-chamber implantable pacemaker, rate-responsive			E2403	No Clinical Signs, Symptoms or Conditions	A071102	Failure to Interrogate		
							F26	No Health Consequences or Impact	A071102	Failure to Interrogate		
18432	2026	All other reportable incidents	47265	Dual-chamber implantable pacemaker, rate-responsive			E0112	Dizziness	A0712	Pacing Problem		
							E060106	Heart Block	A0712	Pacing Problem		
							F2307	Reprogramming of Device	A0712	Pacing Problem		
18433	2026	All other reportable incidents	47265	Dual-chamber implantable pacemaker, rate-responsive			E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
18434	2026	All other reportable incidents	47265	Dual-chamber implantable pacemaker, rate-responsive			E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
18435	2026	All other reportable incidents	47265	Dual-chamber implantable pacemaker, rate-responsive			E2403	No Clinical Signs, Symptoms or Conditions	A070909	Oversensing		
									A090801	Signal Artifact/Noise		
									A1301	Failure to Read Input Signal		
							F26	No Health Consequences or Impact	A070909	Oversensing		
									A090801	Signal Artifact/Noise		
									A1301	Failure to Read Input Signal		

18436	2026	All other reportable incidents	47265	Dual-chamber implantable pacemaker, rate-responsive			E2403	No Clinical Signs, Symptoms or Conditions	A0711	Interrogation Problem		
									A1107	Data Problem		
									A23	Use of Device Problem		
							F26	No Health Consequences or Impact	A0711	Interrogation Problem		
									A1107	Data Problem		
									A23	Use of Device Problem		
18437	2026		17882	Physiologic-monitoring defibrillation system	CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A0201	Product Quality Problem		
									A072001	Electromagnetic Interference		
									A090805	Incorrect Measurement		
							F05	Delay to Treatment/ Therapy	A0201	Product Quality Problem		
									A072001	Electromagnetic Interference		
									A090805	Incorrect Measurement		
18438	2026	All other reportable incidents	47265	Dual-chamber implantable pacemaker, rate-responsive			E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
18439	2026		47267	Single-chamber implantable pacemaker, rate-responsive	CT1685	Cardiac pacemakers	E020201	Anxiety	A27	Appropriate Device Problem Term/Code not Available		
							F1901	Additional Surgery	A27	Appropriate Device Problem Term/Code not Available		
					CT257	Implantable cardiac pacemakers	E020201	Anxiety	A27	Appropriate Device Problem Term/Code not Available		
							F1901	Additional Surgery	A27	Appropriate Device Problem Term/Code not Available		
					CT2866	Implantable cardiac electrical stimulators	E020201	Anxiety	A27	Appropriate Device Problem Term/Code not Available		
							F1901	Additional Surgery	A27	Appropriate Device Problem Term/Code not Available		
18440	2026		47263	Cardiac resynchronization therapy implantable pacemaker	CT1685	Cardiac pacemakers	E020201	Anxiety	A2205	Device Provides Uninformative Message		
							F1901	Additional Surgery	A2205	Device Provides Uninformative Message		
					CT257	Implantable cardiac pacemakers	E020201	Anxiety	A2205	Device Provides Uninformative Message		
							F1901	Additional Surgery	A2205	Device Provides Uninformative Message		
					CT2866	Implantable cardiac electrical stimulators	E020201	Anxiety	A2205	Device Provides Uninformative Message		
							F1901	Additional Surgery	A2205	Device Provides Uninformative Message		
18441	2026	All other reportable incidents	47265	Dual-chamber implantable pacemaker, rate-responsive			E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
18442	2026	All other reportable incidents	47265	Dual-chamber implantable pacemaker, rate-responsive	CT1685	Cardiac pacemakers	E020201	Anxiety	A070504	Premature Discharge of Battery		
									A11	Computer Software Problem		
							F1901	Additional Surgery	A070504	Premature Discharge of Battery		
									A11	Computer Software Problem		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
									A11	Computer Software Problem		
					CT257	Implantable cardiac pacemakers	E020201	Anxiety	A070504	Premature Discharge of Battery		
									A11	Computer Software Problem		
							F1901	Additional Surgery	A070504	Premature Discharge of Battery		
									A11	Computer Software Problem		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
									A11	Computer Software Problem		
					CT2866	Implantable cardiac electrical stimulators	E020201	Anxiety	A070504	Premature Discharge of Battery		
									A11	Computer Software Problem		
							F1901	Additional Surgery	A070504	Premature Discharge of Battery		
									A11	Computer Software Problem		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
									A11	Computer Software Problem		
18443	2026	All other reportable incidents	47265	Dual-chamber implantable pacemaker, rate-responsive			E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
18444	2026		47265	Dual-chamber implantable pacemaker, rate-responsive	CT1685	Cardiac pacemakers	E020201	Anxiety	A2205	Device Provides Uninformative Message		
							F1901	Additional Surgery	A2205	Device Provides Uninformative Message		
					CT257	Implantable cardiac pacemakers	E020201	Anxiety	A2205	Device Provides Uninformative Message		
							F1901	Additional Surgery	A2205	Device Provides Uninformative Message		
					CT2866	Implantable cardiac electrical stimulators	E020201	Anxiety	A2205	Device Provides Uninformative Message		
							F1901	Additional Surgery	A2205	Device Provides Uninformative Message		
18445	2026	All other reportable incidents	47265	Dual-chamber implantable pacemaker, rate-responsive	CT1685	Cardiac pacemakers	E020201	Anxiety				
							F26	No Health Consequences or Impact				
					CT257	Implantable cardiac pacemakers	E020201	Anxiety				
							F26	No Health Consequences or Impact				
					CT2866	Implantable cardiac electrical stimulators	E020201	Anxiety				
							F26	No Health Consequences or Impact				
18446	2026		47267	Single-chamber implantable pacemaker, rate-responsive	CT2750	Implantable cardiac electrical stimulators and associated devices	E020201	Anxiety	A27	Appropriate Device Problem Term/Code not Available		
							F1901	Additional Surgery	A27	Appropriate Device Problem Term/Code not Available		
18447	2026		47265	Dual-chamber implantable pacemaker, rate-responsive	CT1685	Cardiac pacemakers	E020201	Anxiety	A27	Appropriate Device Problem Term/Code not Available		
							F1901	Additional Surgery	A27	Appropriate Device Problem Term/Code not Available		
					CT257	Implantable cardiac pacemakers	E020201	Anxiety	A27	Appropriate Device Problem Term/Code not Available		
							F1901	Additional Surgery	A27	Appropriate Device Problem Term/Code not Available		
					CT2866	Implantable cardiac electrical stimulators	E020201	Anxiety	A27	Appropriate Device Problem Term/Code not Available		
							F1901	Additional Surgery	A27	Appropriate Device Problem Term/Code not Available		
18448	2026		47263	Cardiac resynchronization therapy implantable pacemaker	CT1685	Cardiac pacemakers	E020201	Anxiety	A071205	Pacemaker Found in Back-Up Mode		
									A1606	Premature Indicator Activation		
							F1901	Additional Surgery	A071205	Pacemaker Found in Back-Up Mode		
									A1606	Premature Indicator Activation		
					CT257	Implantable cardiac pacemakers	E020201	Anxiety	A071205	Pacemaker Found in Back-Up Mode		
									A1606	Premature Indicator Activation		
							F1901	Additional Surgery	A071205	Pacemaker Found in Back-Up Mode		
									A1606	Premature Indicator Activation		
					CT2866	Implantable cardiac electrical stimulators	E020201	Anxiety	A071205	Pacemaker Found in Back-Up Mode		

									A1606	Premature Indicator Activation		
							F1901	Additional Surgery	A071205	Pacemaker Found in Back-Up Mode		
									A1606	Premature Indicator Activation		
18449	2026		47265	Dual-chamber implantable pacemaker, rate-responsive	CT2750	Implantable cardiac electrical stimulators and associated devices	E020201	Anxiety	A2205	Device Provides Uninformative Message		
							F1901	Additional Surgery	A2205	Device Provides Uninformative Message		
18450	2026		47265	Dual-chamber implantable pacemaker, rate-responsive	CT1685	Cardiac pacemakers	E020201	Anxiety	A071205	Pacemaker Found in Back-Up Mode		
									A2205	Device Provides Uninformative Message		
									A27	Appropriate Device Problem Term/Code not Available		
							F1901	Additional Surgery	A071205	Pacemaker Found in Back-Up Mode		
									A2205	Device Provides Uninformative Message		
									A27	Appropriate Device Problem Term/Code not Available		
					CT257	Implantable cardiac pacemakers	E020201	Anxiety	A071205	Pacemaker Found in Back-Up Mode		
									A2205	Device Provides Uninformative Message		
									A27	Appropriate Device Problem Term/Code not Available		
							F1901	Additional Surgery	A071205	Pacemaker Found in Back-Up Mode		
									A2205	Device Provides Uninformative Message		
									A27	Appropriate Device Problem Term/Code not Available		
					CT2866	Implantable cardiac electrical stimulators	E020201	Anxiety	A071205	Pacemaker Found in Back-Up Mode		
									A2205	Device Provides Uninformative Message		
									A27	Appropriate Device Problem Term/Code not Available		
							F1901	Additional Surgery	A071205	Pacemaker Found in Back-Up Mode		
									A2205	Device Provides Uninformative Message		
									A27	Appropriate Device Problem Term/Code not Available		
18451	2026	All other reportable incidents	47265	Dual-chamber implantable pacemaker, rate-responsive			E2403	No Clinical Signs, Symptoms or Conditions	A0908	Incorrect, Inadequate or Imprecise Result or Readings		
							F26	No Health Consequences or Impact	A0908	Incorrect, Inadequate or Imprecise Result or Readings		
18452	2026	All other reportable incidents	47265	Dual-chamber implantable pacemaker, rate-responsive			E0717	Dyspnea	A26	Insufficient Device Problem Information		
							F11	Non-Serious Injury/ Illness/ Impairment	A26	Insufficient Device Problem Information		
18453	2026	All other reportable incidents	47265	Dual-chamber implantable pacemaker, rate-responsive			E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
18454	2026	All other reportable incidents	47265	Dual-chamber implantable pacemaker, rate-responsive			E2403	No Clinical Signs, Symptoms or Conditions	A0908	Incorrect, Inadequate or Imprecise Result or Readings		
							F26	No Health Consequences or Impact	A0908	Incorrect, Inadequate or Imprecise Result or Readings		
18455	2026	All other reportable incidents	47265	Dual-chamber implantable pacemaker, rate-responsive	CT1685	Cardiac pacemakers	E0112	Dizziness	A26	Insufficient Device Problem Information		
							E013403	Paresthesia	A26	Insufficient Device Problem Information		
							E020201	Anxiety	A26	Insufficient Device Problem Information		
							F11	Non-Serious Injury/ Illness/ Impairment	A26	Insufficient Device Problem Information		
							E0629	Palpitations	A26	Insufficient Device Problem Information		
					CT257	Implantable cardiac pacemakers	E0112	Dizziness	A26	Insufficient Device Problem Information		
							E013403	Paresthesia	A26	Insufficient Device Problem Information		
							E020201	Anxiety	A26	Insufficient Device Problem Information		
							F11	Non-Serious Injury/ Illness/ Impairment	A26	Insufficient Device Problem Information		
							E0629	Palpitations	A26	Insufficient Device Problem Information		
					CT2866	Implantable cardiac electrical stimulators	E0112	Dizziness	A26	Insufficient Device Problem Information		
							E013403	Paresthesia	A26	Insufficient Device Problem Information		
							E020201	Anxiety	A26	Insufficient Device Problem Information		
							F11	Non-Serious Injury/ Illness/ Impairment	A26	Insufficient Device Problem Information		
							E0629	Palpitations	A26	Insufficient Device Problem Information		
18456	2026	All other reportable incidents	47265	Dual-chamber implantable pacemaker, rate-responsive			E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
18457	2026	All other reportable incidents	60789	Intracardiac pacemaker			E0503	Embolism/Embolus	A051201	Device Dislodged or Dislocated	15/04/2026	15/04/2026
							E0514	Thrombosis/Thrombus	A051201	Device Dislodged or Dislocated	15/04/2026	15/04/2026
							F1908	Prolonged Surgery	A051201	Device Dislodged or Dislocated	15/04/2026	15/04/2026
18458	2026	All other reportable incidents	47265	Dual-chamber implantable pacemaker, rate-responsive			E2403	No Clinical Signs, Symptoms or Conditions	A070101	Failure to Capture		
									A070909	Oversensing		
							F1010	Inadequate Treatment	A070101	Failure to Capture		
									A070909	Oversensing		
18459	2026	All other reportable incidents	47265	Dual-chamber implantable pacemaker, rate-responsive			E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
18460	2026		35852	Single-chamber implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A0201	Product Quality Problem		
									A070504	Premature Discharge of Battery		
									A160601	Premature Elective Replacement Indicator		
							F0101	Therapeutic Response Decreased	A0201	Product Quality Problem		
									A070504	Premature Discharge of Battery		
									A160601	Premature Elective Replacement Indicator		
18461	2026	All other reportable incidents	47265	Dual-chamber implantable pacemaker, rate-responsive			E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
18462	2026	All other reportable incidents	47265	Dual-chamber implantable pacemaker, rate-responsive			E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
18463	2026	All other reportable incidents	47267	Single-chamber implantable pacemaker, rate-responsive			E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
18464	2026	All other reportable incidents	47265	Dual-chamber implantable pacemaker, rate-responsive			E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	26/04/2026	26/04/2026
							F1901	Additional Surgery	A070504	Premature Discharge of Battery	26/04/2026	26/04/2026
18465	2026	All other reportable incidents	47265	Dual-chamber implantable pacemaker, rate-responsive			E2403	No Clinical Signs, Symptoms or Conditions	A0711	Interrogation Problem		
							F26	No Health Consequences or Impact	A0711	Interrogation Problem		
18466	2026	All other reportable incidents	47265	Dual-chamber implantable pacemaker, rate-responsive			E0109	Convulsion/Seizure	A26	Insufficient Device Problem Information		
							E0112	Dizziness	A26	Insufficient Device Problem Information		
							E011903	Syncope/Fainting	A26	Insufficient Device Problem Information		
							F12	Serious Injury/ Illness/ Impairment	A26	Insufficient Device Problem Information		
18467	2026	All other reportable incidents	47265	Dual-chamber implantable pacemaker, rate-responsive			E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
18468	2026	All other reportable incidents	47265	Dual-chamber implantable pacemaker, rate-			E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		

				responsive			F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
18469	2026	All other reportable incidents	47265	Dual-chamber implantable pacemaker, rate-responsive			E2403	No Clinical Signs, Symptoms or Conditions	A070101	Failure to Capture		
							F1010	Inadequate Treatment	A070101	Failure to Capture		
18470	2026	All other reportable incidents	47265	Dual-chamber implantable pacemaker, rate-responsive			E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
18471	2026	All other reportable incidents	47265	Dual-chamber implantable pacemaker, rate-responsive			E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
18472	2026	All other reportable incidents	47267	Single-chamber implantable pacemaker, rate-responsive			E020201	Anxiety	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
18473	2026	All other reportable incidents	47265	Dual-chamber implantable pacemaker, rate-responsive			E020201	Anxiety	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
18474	2026	All other reportable incidents	47265	Dual-chamber implantable pacemaker, rate-responsive			E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
18475	2026	All other reportable incidents	47265	Dual-chamber implantable pacemaker, rate-responsive			E2403	No Clinical Signs, Symptoms or Conditions	A071102	Failure to Interrogate		
							F26	No Health Consequences or Impact	A071102	Failure to Interrogate		
18476	2026	All other reportable incidents	47265	Dual-chamber implantable pacemaker, rate-responsive			E2403	No Clinical Signs, Symptoms or Conditions	A071102	Failure to Interrogate		
							F26	No Health Consequences or Impact	A071102	Failure to Interrogate		
18477	2026	All other reportable incidents	47265	Dual-chamber implantable pacemaker, rate-responsive	CT1685	Cardiac pacemakers	E020201	Anxiety	A070504	Premature Discharge of Battery		
									A27	Appropriate Device Problem Term/Code not Available		
							F1901	Additional Surgery	A070504	Premature Discharge of Battery		
									A27	Appropriate Device Problem Term/Code not Available		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
									A27	Appropriate Device Problem Term/Code not Available		
					CT257	Implantable cardiac pacemakers	E020201	Anxiety	A070504	Premature Discharge of Battery		
									A27	Appropriate Device Problem Term/Code not Available		
							F1901	Additional Surgery	A070504	Premature Discharge of Battery		
									A27	Appropriate Device Problem Term/Code not Available		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
									A27	Appropriate Device Problem Term/Code not Available		
					CT2866	Implantable cardiac electrical stimulators	E020201	Anxiety	A070504	Premature Discharge of Battery		
									A27	Appropriate Device Problem Term/Code not Available		
							F1901	Additional Surgery	A070504	Premature Discharge of Battery		
									A27	Appropriate Device Problem Term/Code not Available		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
									A27	Appropriate Device Problem Term/Code not Available		
18478	2026	All other reportable incidents	60789	Intracardiac pacemaker			E0503	Embolism/Embolus	A051201	Device Dislodged or Dislocated	27/04/2026	27/04/2026
									A070101	Failure to Capture	27/04/2026	27/04/2026
									A070908	Failure to Sense	27/04/2026	27/04/2026
							F1901	Additional Surgery	A051201	Device Dislodged or Dislocated	27/04/2026	27/04/2026
									A070101	Failure to Capture	27/04/2026	27/04/2026
									A070908	Failure to Sense	27/04/2026	27/04/2026
18479	2026	All other reportable incidents	47265	Dual-chamber implantable pacemaker, rate-responsive			E011903	Syncope/Fainting	A26	Insufficient Device Problem Information		
							E0717	Dyspnea	A26	Insufficient Device Problem Information		
							E233001	Chest Pain	A26	Insufficient Device Problem Information		
							F08	Hospitalization or Prolonged Hospitalization	A26	Insufficient Device Problem Information		
18480	2026	Death	47265	Dual-chamber implantable pacemaker, rate-responsive	CT1685	Cardiac pacemakers	E0602	Cardiac Arrest	A24	No Apparent Device Problem		
							F02	Death	A24	No Apparent Device Problem		
					CT257	Implantable cardiac pacemakers	E0602	Cardiac Arrest	A24	No Apparent Device Problem		
							F02	Death	A24	No Apparent Device Problem		
					CT2866	Implantable cardiac electrical stimulators	E0602	Cardiac Arrest	A24	No Apparent Device Problem		
							F02	Death	A24	No Apparent Device Problem		
18481	2026		47265	Dual-chamber implantable pacemaker, rate-responsive	CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A0203	Defective Device		
							F1903	Device Explantation	A0203	Defective Device		
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A0203	Defective Device		
							F1903	Device Explantation	A0203	Defective Device		
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A0203	Defective Device		
							F1903	Device Explantation	A0203	Defective Device		
18482	2026	All other reportable incidents	60789	Intracardiac pacemaker			E2403	No Clinical Signs, Symptoms or Conditions	A070102	High Capture Threshold		
									A0705	Battery Problem		
							F26	No Health Consequences or Impact	A070102	High Capture Threshold		
									A0705	Battery Problem		
18483	2026	All other reportable incidents	47265	Dual-chamber implantable pacemaker, rate-responsive			E2403	No Clinical Signs, Symptoms or Conditions	A0908	Incorrect, Inadequate or Imprecise Result or Readings	30/04/2026	30/04/2026
							F08	Hospitalization or Prolonged Hospitalization	A0908	Incorrect, Inadequate or Imprecise Result or Readings	30/04/2026	30/04/2026
							F1905	Device Revision or Replacement	A0908	Incorrect, Inadequate or Imprecise Result or Readings	30/04/2026	30/04/2026
18484	2026	All other reportable incidents	47267	Single-chamber implantable pacemaker, rate-responsive			E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem		
							F26	No Health Consequences or Impact	A0705	Battery Problem		
18485	2026	All other reportable incidents	47267	Single-chamber implantable pacemaker, rate-responsive			E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
18486	2026	All other reportable incidents	47265	Dual-chamber implantable pacemaker, rate-responsive			E020201	Anxiety	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
18487	2026	All other reportable incidents	47265	Dual-chamber implantable pacemaker, rate-responsive			E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
18488	2026	All other reportable incidents	47265	Dual-chamber implantable pacemaker, rate-responsive			E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
18489	2026	All other reportable incidents	47265	Dual-chamber implantable pacemaker, rate-responsive			E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
18490	2026	All other reportable incidents	47265	Dual-chamber implantable pacemaker, rate-responsive			E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
18491	2026	All other reportable incidents	60789	Intracardiac pacemaker			E0604	Cardiac Perforation	A24	No Apparent Device Problem	29/04/2026	29/04/2026

							E0605	Cardiac Tamponade	A24	No Apparent Device Problem	29/04/2026	29/04/2026
							E0619	Pericardial Effusion	A24	No Apparent Device Problem	29/04/2026	29/04/2026
							E0742	Respiratory Failure	A24	No Apparent Device Problem	29/04/2026	29/04/2026
							F1203	Life Threatening Illness or Injury	A24	No Apparent Device Problem	29/04/2026	29/04/2026
18492	2026	All other reportable incidents	47265	Dual-chamber implantable pacemaker, rate-responsive			E011903	Syncope/Fainting	A070909	Oversensing	01/05/2026	01/05/2026
									A0712	Pacing Problem	01/05/2026	01/05/2026
									A0908	Incorrect, Inadequate or Imprecise Result or Readings	01/05/2026	01/05/2026
									A090801	Signal Artifact/Noise	01/05/2026	01/05/2026
							F08	Hospitalization or Prolonged Hospitalization	A070909	Oversensing	01/05/2026	01/05/2026
									A0712	Pacing Problem	01/05/2026	01/05/2026
									A0908	Incorrect, Inadequate or Imprecise Result or Readings	01/05/2026	01/05/2026
									A090801	Signal Artifact/Noise	01/05/2026	01/05/2026
							F1905	Device Revision or Replacement	A070909	Oversensing	01/05/2026	01/05/2026
									A0712	Pacing Problem	01/05/2026	01/05/2026
									A0908	Incorrect, Inadequate or Imprecise Result or Readings	01/05/2026	01/05/2026
									A090801	Signal Artifact/Noise	01/05/2026	01/05/2026
							F1010	Inadequate Treatment	A070909	Oversensing	01/05/2026	01/05/2026
									A0712	Pacing Problem	01/05/2026	01/05/2026
									A0908	Incorrect, Inadequate or Imprecise Result or Readings	01/05/2026	01/05/2026
									A090801	Signal Artifact/Noise	01/05/2026	01/05/2026
18493	2026	All other reportable incidents	47265	Dual-chamber implantable pacemaker, rate-responsive			E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
18494	2026	All other reportable incidents	47265	Dual-chamber implantable pacemaker, rate-responsive			E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
18495	2026	All other reportable incidents	47265	Dual-chamber implantable pacemaker, rate-responsive			E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
18496	2026	All other reportable incidents	47265	Dual-chamber implantable pacemaker, rate-responsive			E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
18497	2026	All other reportable incidents	47265	Dual-chamber implantable pacemaker, rate-responsive			E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
18498	2026	All other reportable incidents	47265	Dual-chamber implantable pacemaker, rate-responsive			E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
18499	2026	All other reportable incidents	47265	Dual-chamber implantable pacemaker, rate-responsive			E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
18500	2026	All other reportable incidents	47265	Dual-chamber implantable pacemaker, rate-responsive			E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
18501	2026	All other reportable incidents	47265	Dual-chamber implantable pacemaker, rate-responsive			E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
18502	2026	All other reportable incidents	47265	Dual-chamber implantable pacemaker, rate-responsive			E2403	No Clinical Signs, Symptoms or Conditions	A0908	Incorrect, Inadequate or Imprecise Result or Readings	24/04/2026	24/04/2026
							F08	Hospitalization or Prolonged Hospitalization	A0908	Incorrect, Inadequate or Imprecise Result or Readings	24/04/2026	24/04/2026
							F1905	Device Revision or Replacement	A0908	Incorrect, Inadequate or Imprecise Result or Readings	24/04/2026	24/04/2026
18503	2026	All other reportable incidents	47265	Dual-chamber implantable pacemaker, rate-responsive			E0112	Dizziness	A0908	Incorrect, Inadequate or Imprecise Result or Readings		
							E020204	Malaise	A0908	Incorrect, Inadequate or Imprecise Result or Readings		
							F11	Non-Serious Injury/ Illness/ Impairment	A0908	Incorrect, Inadequate or Imprecise Result or Readings		
18504	2026	All other reportable incidents	47265	Dual-chamber implantable pacemaker, rate-responsive			E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
18505	2026	All other reportable incidents	47265	Dual-chamber implantable pacemaker, rate-responsive			E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
18506	2026	All other reportable incidents	47265	Dual-chamber implantable pacemaker, rate-responsive			E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
18507	2026	All other reportable incidents	47265	Dual-chamber implantable pacemaker, rate-responsive			E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery		
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery		
18508	2026	All other reportable incidents	47265	Dual-chamber implantable pacemaker, rate-responsive			E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
18509	2026	All other reportable incidents	47267	Single-chamber implantable pacemaker, rate-responsive			E2403	No Clinical Signs, Symptoms or Conditions	A070909	Oversensing		
									A070910	Undersensing		
							F26	No Health Consequences or Impact	A070909	Oversensing		
									A070910	Undersensing		
18510	2026	Serious public health threat	17882	Physiologic-monitoring defibrillation system	CT1684	Cardiac pacemakers and associated devices	E0623	Unspecified Heart Problem	A11	Computer Software Problem		
							F05	Delay to Treatment/ Therapy	A11	Computer Software Problem		
18511	2026	All other reportable incidents	47265	Dual-chamber implantable pacemaker, rate-responsive			E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
18512	2026	All other reportable incidents	47265	Dual-chamber implantable pacemaker, rate-responsive			E2403	No Clinical Signs, Symptoms or Conditions	A0908	Incorrect, Inadequate or Imprecise Result or Readings		
							F26	No Health Consequences or Impact	A0908	Incorrect, Inadequate or Imprecise Result or Readings		
18513	2026	All other reportable incidents	47267	Single-chamber implantable pacemaker, rate-responsive			E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
18514	2026	All other reportable incidents	47265	Dual-chamber implantable pacemaker, rate-responsive			E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
18515	2026	All other reportable incidents	47265	Dual-chamber implantable pacemaker, rate-responsive			E2403	No Clinical Signs, Symptoms or Conditions	A160105	False Alarm		
							F26	No Health Consequences or Impact	A160105	False Alarm		
18516	2026	All other reportable incidents	47265	Dual-chamber implantable pacemaker, rate-responsive			E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F1903	Device Explantation	A070504	Premature Discharge of Battery		
18517	2026	All other reportable incidents	47265	Dual-chamber implantable pacemaker, rate-responsive			E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
18518	2026	All other reportable incidents	47267	Single-chamber implantable pacemaker, rate-responsive			E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
18519	2026	All other reportable incidents	47265	Dual-chamber implantable pacemaker, rate-responsive			E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		

18520	2026	All other reportable incidents	47265	Dual-chamber implantable pacemaker, rate-responsive			E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
18521	2026	All other reportable incidents	47265	Dual-chamber implantable pacemaker, rate-responsive			E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
18522	2026	All other reportable incidents	47265	Dual-chamber implantable pacemaker, rate-responsive			E2403	No Clinical Signs, Symptoms or Conditions	A070101	Failure to Capture		
							F1010	Inadequate Treatment	A070101	Failure to Capture		
18523	2026	All other reportable incidents	47265	Dual-chamber implantable pacemaker, rate-responsive	CT1685	Cardiac pacemakers	E011903	Syncope/Fainting	A0712	Pacing Problem		
									A23	Use of Device Problem		
							E2007	Fall	A0712	Pacing Problem		
									A23	Use of Device Problem		
							F12	Serious Injury/ Illness/ Impairment	A0712	Pacing Problem		
									A23	Use of Device Problem		
							F2307	Reprogramming of Device	A0712	Pacing Problem		
									A23	Use of Device Problem		
					CT257	Implantable cardiac pacemakers	E011903	Syncope/Fainting	A0712	Pacing Problem		
									A23	Use of Device Problem		
							E2007	Fall	A0712	Pacing Problem		
									A23	Use of Device Problem		
							F12	Serious Injury/ Illness/ Impairment	A0712	Pacing Problem		
									A23	Use of Device Problem		
							F2307	Reprogramming of Device	A0712	Pacing Problem		
									A23	Use of Device Problem		
					CT2866	Implantable cardiac electrical stimulators	E011903	Syncope/Fainting	A0712	Pacing Problem		
									A23	Use of Device Problem		
							E2007	Fall	A0712	Pacing Problem		
									A23	Use of Device Problem		
							F12	Serious Injury/ Illness/ Impairment	A0712	Pacing Problem		
									A23	Use of Device Problem		
							F2307	Reprogramming of Device	A0712	Pacing Problem		
									A23	Use of Device Problem		
18524	2026	All other reportable incidents	47265	Dual-chamber implantable pacemaker, rate-responsive			E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
18525	2026	All other reportable incidents	47265	Dual-chamber implantable pacemaker, rate-responsive			E2403	No Clinical Signs, Symptoms or Conditions	A070101	Failure to Capture		
									A070102	High Capture Threshold		
									A070909	Oversensing		
									A090801	Signal Artifact/Noise		
									A1301	Failure to Read Input Signal		
							F1010	Inadequate Treatment	A070101	Failure to Capture		
									A070102	High Capture Threshold		
									A070909	Oversensing		
									A090801	Signal Artifact/Noise		
									A1301	Failure to Read Input Signal		
18526	2026	All other reportable incidents	47265	Dual-chamber implantable pacemaker, rate-responsive			E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
18527	2026	All other reportable incidents	47265	Dual-chamber implantable pacemaker, rate-responsive			E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
18528	2026	All other reportable incidents	47265	Dual-chamber implantable pacemaker, rate-responsive			E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
18529	2026	All other reportable incidents	47265	Dual-chamber implantable pacemaker, rate-responsive			E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
18530	2026	All other reportable incidents	47265	Dual-chamber implantable pacemaker, rate-responsive			E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
18531	2026	All other reportable incidents	47265	Dual-chamber implantable pacemaker, rate-responsive			E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
18532	2026	All other reportable incidents	47265	Dual-chamber implantable pacemaker, rate-responsive			E2403	No Clinical Signs, Symptoms or Conditions	A070910	Undersensing		
							F26	No Health Consequences or Impact	A070910	Undersensing		
18533	2026	All other reportable incidents	35852	Single-chamber implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A070909	Oversensing		
							F26	No Health Consequences or Impact	A070909	Oversensing		
18534	2026		47265	Dual-chamber implantable pacemaker, rate-responsive	CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
									A0722	Impedance Problem		
									A0908	Incorrect, Inadequate or Imprecise Result or Readings		
							F0102	Therapeutic Response Increased	A070504	Premature Discharge of Battery		
									A0722	Impedance Problem		
									A0908	Incorrect, Inadequate or Imprecise Result or Readings		
							F28	Appropriate Health Impact Term/Code not Available	A070504	Premature Discharge of Battery		
									A0722	Impedance Problem		
									A0908	Incorrect, Inadequate or Imprecise Result or Readings		
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
									A0722	Impedance Problem		
									A0908	Incorrect, Inadequate or Imprecise Result or Readings		
							F0102	Therapeutic Response Increased	A070504	Premature Discharge of Battery		
									A0722	Impedance Problem		
									A0908	Incorrect, Inadequate or Imprecise Result or Readings		
							F28	Appropriate Health Impact Term/Code not Available	A070504	Premature Discharge of Battery		
									A0722	Impedance Problem		
									A0908	Incorrect, Inadequate or Imprecise Result or Readings		
					CT2750	Implantable cardiac electrical stimulators and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
									A0722	Impedance Problem		
									A0908	Incorrect, Inadequate or Imprecise Result or Readings		
							F0102	Therapeutic Response Increased	A070504	Premature Discharge of Battery		
									A0722	Impedance Problem		
									A0908	Incorrect, Inadequate or Imprecise Result or Readings		

							F28	Appropriate Health Impact Term/Code not Available	A070504	Premature Discharge of Battery		
									A0722	Impedance Problem		
									A0908	Incorrect, Inadequate or Imprecise Result or Readings		
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
									A0722	Impedance Problem		
									A0908	Incorrect, Inadequate or Imprecise Result or Readings		
							F0102	Therapeutic Response Increased	A070504	Premature Discharge of Battery		
									A0722	Impedance Problem		
									A0908	Incorrect, Inadequate or Imprecise Result or Readings		
							F28	Appropriate Health Impact Term/Code not Available	A070504	Premature Discharge of Battery		
									A0722	Impedance Problem		
									A0908	Incorrect, Inadequate or Imprecise Result or Readings		
18535	2026		47265	Dual-chamber implantable pacemaker, rate-responsive	CT1685	Cardiac pacemakers	E020201	Anxiety	A071205	Pacemaker Found in Back-Up Mode		
									A27	Appropriate Device Problem Term/Code not Available		
							F1901	Additional Surgery	A071205	Pacemaker Found in Back-Up Mode		
									A27	Appropriate Device Problem Term/Code not Available		
					CT257	Implantable cardiac pacemakers	E020201	Anxiety	A071205	Pacemaker Found in Back-Up Mode		
									A27	Appropriate Device Problem Term/Code not Available		
							F1901	Additional Surgery	A071205	Pacemaker Found in Back-Up Mode		
									A27	Appropriate Device Problem Term/Code not Available		
					CT2750	Implantable cardiac electrical stimulators and associated devices	E020201	Anxiety	A071205	Pacemaker Found in Back-Up Mode		
									A27	Appropriate Device Problem Term/Code not Available		
							F1901	Additional Surgery	A071205	Pacemaker Found in Back-Up Mode		
									A27	Appropriate Device Problem Term/Code not Available		
					CT2866	Implantable cardiac electrical stimulators	E020201	Anxiety	A071205	Pacemaker Found in Back-Up Mode		
									A27	Appropriate Device Problem Term/Code not Available		
							F1901	Additional Surgery	A071205	Pacemaker Found in Back-Up Mode		
									A27	Appropriate Device Problem Term/Code not Available		
18536	2026	All other reportable incidents	47265	Dual-chamber implantable pacemaker, rate-responsive			E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
18537	2026	All other reportable incidents	47265	Dual-chamber implantable pacemaker, rate-responsive			E2403	No Clinical Signs, Symptoms or Conditions	A070909	Oversensing		
									A090801	Signal Artifact/Noise		
							F2203	Imaging Required	A070909	Oversensing		
									A090801	Signal Artifact/Noise		
							F26	No Health Consequences or Impact	A070909	Oversensing		
									A090801	Signal Artifact/Noise		
18538	2026	All other reportable incidents	47265	Dual-chamber implantable pacemaker, rate-responsive			E0112	Dizziness	A070101	Failure to Capture	14/05/2026	14/05/2026
									A070504	Premature Discharge of Battery	14/05/2026	14/05/2026
									A071102	Failure to Interrogate	14/05/2026	14/05/2026
							F1901	Additional Surgery	A070101	Failure to Capture	14/05/2026	14/05/2026
									A070504	Premature Discharge of Battery	14/05/2026	14/05/2026
									A071102	Failure to Interrogate	14/05/2026	14/05/2026
18539	2026	All other reportable incidents	47265	Dual-chamber implantable pacemaker, rate-responsive			E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
18540	2026	All other reportable incidents	47265	Dual-chamber implantable pacemaker, rate-responsive			E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
18541	2026	All other reportable incidents	47265	Dual-chamber implantable pacemaker, rate-responsive			E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
18542	2026	All other reportable incidents	47265	Dual-chamber implantable pacemaker, rate-responsive	CT1685	Cardiac pacemakers	E0112	Dizziness	A27	Appropriate Device Problem Term/Code not Available		
							E011903	Syncope/Fainting	A27	Appropriate Device Problem Term/Code not Available		
							E060109	Tachycardia	A27	Appropriate Device Problem Term/Code not Available		
							F1203	Life Threatening Illness or Injury	A27	Appropriate Device Problem Term/Code not Available		
					CT257	Implantable cardiac pacemakers	E0112	Dizziness	A27	Appropriate Device Problem Term/Code not Available		
							E011903	Syncope/Fainting	A27	Appropriate Device Problem Term/Code not Available		
							E060109	Tachycardia	A27	Appropriate Device Problem Term/Code not Available		
							F1203	Life Threatening Illness or Injury	A27	Appropriate Device Problem Term/Code not Available		
					CT2866	Implantable cardiac electrical stimulators	E0112	Dizziness	A27	Appropriate Device Problem Term/Code not Available		
							E011903	Syncope/Fainting	A27	Appropriate Device Problem Term/Code not Available		
							E060109	Tachycardia	A27	Appropriate Device Problem Term/Code not Available		
							F1203	Life Threatening Illness or Injury	A27	Appropriate Device Problem Term/Code not Available		
18543	2026	All other reportable incidents	47265	Dual-chamber implantable pacemaker, rate-responsive			E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	09/03/2026	09/03/2026
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	09/03/2026	09/03/2026
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	09/03/2026	09/03/2026
18544	2026	All other reportable incidents	47265	Dual-chamber implantable pacemaker, rate-responsive			E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
18545	2026	All other reportable incidents	47265	Dual-chamber implantable pacemaker, rate-responsive			E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
18546	2026	All other reportable incidents	47267	Single-chamber implantable pacemaker, rate-responsive			E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		

18547	2026	All other reportable incidents	47265	Dual-chamber implantable pacemaker, rate-responsive			E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	01/05/2026	01/05/2026
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	01/05/2026	01/05/2026
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	01/05/2026	01/05/2026
18548	2026	All other reportable incidents	47265	Dual-chamber implantable pacemaker, rate-responsive			E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
18549	2026	All other reportable incidents	47265	Dual-chamber implantable pacemaker, rate-responsive			E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
18550	2026		47265	Dual-chamber implantable pacemaker, rate-responsive	CT1685	Cardiac pacemakers	E020201	Anxiety	A071205	Pacemaker Found in Back-Up Mode		
									A27	Appropriate Device Problem Term/Code not Available		
							F1901	Additional Surgery	A071205	Pacemaker Found in Back-Up Mode		
									A27	Appropriate Device Problem Term/Code not Available		
					CT257	Implantable cardiac pacemakers	E020201	Anxiety	A071205	Pacemaker Found in Back-Up Mode		
									A27	Appropriate Device Problem Term/Code not Available		
							F1901	Additional Surgery	A071205	Pacemaker Found in Back-Up Mode		
									A27	Appropriate Device Problem Term/Code not Available		
					CT2750	Implantable cardiac electrical stimulators and associated devices	E020201	Anxiety	A071205	Pacemaker Found in Back-Up Mode		
									A27	Appropriate Device Problem Term/Code not Available		
							F1901	Additional Surgery	A071205	Pacemaker Found in Back-Up Mode		
									A27	Appropriate Device Problem Term/Code not Available		
					CT2866	Implantable cardiac electrical stimulators	E020201	Anxiety	A071205	Pacemaker Found in Back-Up Mode		
									A27	Appropriate Device Problem Term/Code not Available		
							F1901	Additional Surgery	A071205	Pacemaker Found in Back-Up Mode		
									A27	Appropriate Device Problem Term/Code not Available		
18551	2026		47265	Dual-chamber implantable pacemaker, rate-responsive	CT1685	Cardiac pacemakers	E020201	Anxiety	A071205	Pacemaker Found in Back-Up Mode		
									A1103	Program or Algorithm Execution Problem		
									A27	Appropriate Device Problem Term/Code not Available		
							F1901	Additional Surgery	A071205	Pacemaker Found in Back-Up Mode		
									A1103	Program or Algorithm Execution Problem		
									A27	Appropriate Device Problem Term/Code not Available		
					CT257	Implantable cardiac pacemakers	E020201	Anxiety	A071205	Pacemaker Found in Back-Up Mode		
									A1103	Program or Algorithm Execution Problem		
									A27	Appropriate Device Problem Term/Code not Available		
							F1901	Additional Surgery	A071205	Pacemaker Found in Back-Up Mode		
									A1103	Program or Algorithm Execution Problem		
									A27	Appropriate Device Problem Term/Code not Available		
					CT2750	Implantable cardiac electrical stimulators and associated devices	E020201	Anxiety	A071205	Pacemaker Found in Back-Up Mode		
									A1103	Program or Algorithm Execution Problem		
									A27	Appropriate Device Problem Term/Code not Available		
							F1901	Additional Surgery	A071205	Pacemaker Found in Back-Up Mode		
									A1103	Program or Algorithm Execution Problem		
									A27	Appropriate Device Problem Term/Code not Available		
					CT2866	Implantable cardiac electrical stimulators	E020201	Anxiety	A071205	Pacemaker Found in Back-Up Mode		
									A1103	Program or Algorithm Execution Problem		
									A27	Appropriate Device Problem Term/Code not Available		
							F1901	Additional Surgery	A071205	Pacemaker Found in Back-Up Mode		
									A1103	Program or Algorithm Execution Problem		
									A27	Appropriate Device Problem Term/Code not Available		
18552	2026	All other reportable incidents	47265	Dual-chamber implantable pacemaker, rate-responsive			E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
18553	2026	All other reportable incidents	47265	Dual-chamber implantable pacemaker, rate-responsive			E2403	No Clinical Signs, Symptoms or Conditions	A070101	Failure to Capture		
									A1301	Failure to Read Input Signal		
							F1010	Inadequate Treatment	A070101	Failure to Capture		
									A1301	Failure to Read Input Signal		
18554	2026	All other reportable incidents	47265	Dual-chamber implantable pacemaker, rate-responsive			E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
18555	2026	All other reportable incidents	47265	Dual-chamber implantable pacemaker, rate-responsive			E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
18556	2026	All other reportable incidents	47265	Dual-chamber implantable pacemaker, rate-responsive			E060104	Bradycardia	A070909	Oversensing		
									A0712	Pacing Problem		
									A072201	High Impedance		
							F1001	Absence of Treatment	A070909	Oversensing		
									A0712	Pacing Problem		
									A072201	High Impedance		
18557	2026	All other reportable incidents	47265	Dual-chamber implantable pacemaker, rate-responsive			E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
18558	2026	All other reportable incidents	47265	Dual-chamber implantable pacemaker, rate-responsive			E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	06/05/2026	06/05/2026
							F1901	Additional Surgery	A070504	Premature Discharge of Battery	06/05/2026	06/05/2026
18559	2026	All other reportable incidents	47265	Dual-chamber implantable pacemaker, rate-responsive			E2403	No Clinical Signs, Symptoms or Conditions	A070909	Oversensing		
									A1301	Failure to Read Input Signal		
							F26	No Health Consequences or Impact	A070909	Oversensing		
									A1301	Failure to Read Input Signal		
18560	2026	All other reportable incidents	47265	Dual-chamber implantable pacemaker, rate-responsive			E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
18561	2026	All other reportable incidents	60789	Intracardiac pacemaker			E0503	Embolism/Embolus	A051201	Device Dislodged or Dislocated	15/05/2026	15/05/2026
									A150302	Difficult or Delayed Separation	15/05/2026	15/05/2026
							F1908	Prolonged Surgery	A051201	Device Dislodged or Dislocated	15/05/2026	15/05/2026

									A150302	Difficult or Delayed Separation	15/05/2026	15/05/2026
18562	2026	Death	47265	Dual-chamber implantable pacemaker, rate-responsive	CT1685	Cardiac pacemakers	E0306	Sepsis			03/12/2024	03/12/2024
							E0610	Endocarditis			03/12/2024	03/12/2024
							E2336	Shock			03/12/2024	03/12/2024
							F02	Death			03/12/2024	03/12/2024
							F08	Hospitalization or Prolonged Hospitalization			03/12/2024	03/12/2024
							F1903	Device Explantation			03/12/2024	03/12/2024
					CT257	Implantable cardiac pacemakers	E0306	Sepsis			03/12/2024	03/12/2024
							E0610	Endocarditis			03/12/2024	03/12/2024
							E2336	Shock			03/12/2024	03/12/2024
							F02	Death			03/12/2024	03/12/2024
							F08	Hospitalization or Prolonged Hospitalization			03/12/2024	03/12/2024
							F1903	Device Explantation			03/12/2024	03/12/2024
					CT2866	Implantable cardiac electrical stimulators	E0306	Sepsis			03/12/2024	03/12/2024
							E0610	Endocarditis			03/12/2024	03/12/2024
							E2336	Shock			03/12/2024	03/12/2024
							F02	Death			03/12/2024	03/12/2024
							F08	Hospitalization or Prolonged Hospitalization			03/12/2024	03/12/2024
							F1903	Device Explantation			03/12/2024	03/12/2024
18563	2026	All other reportable incidents	47265	Dual-chamber implantable pacemaker, rate-responsive			E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
18564	2026	All other reportable incidents	47267	Single-chamber implantable pacemaker, rate-responsive			E060101	Asystole	A070101	Failure to Capture		
							F1010	Inadequate Treatment	A070101	Failure to Capture		
18565	2026	All other reportable incidents	47265	Dual-chamber implantable pacemaker, rate-responsive			E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
18566	2026	All other reportable incidents	47265	Dual-chamber implantable pacemaker, rate-responsive			E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
18567	2026	All other reportable incidents	47265	Dual-chamber implantable pacemaker, rate-responsive			E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
18568	2026	All other reportable incidents	47265	Dual-chamber implantable pacemaker, rate-responsive			E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
18569	2026	All other reportable incidents	47265	Dual-chamber implantable pacemaker, rate-responsive			E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
18570	2026	All other reportable incidents	47265	Dual-chamber implantable pacemaker, rate-responsive			E020201	Anxiety	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
18571	2026	All other reportable incidents	47265	Dual-chamber implantable pacemaker, rate-responsive			E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
18572	2026	All other reportable incidents	47265	Dual-chamber implantable pacemaker, rate-responsive			E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
18573	2026	All other reportable incidents	47265	Dual-chamber implantable pacemaker, rate-responsive			E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
18574	2026	All other reportable incidents	47265	Dual-chamber implantable pacemaker, rate-responsive			E020201	Anxiety	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
18575	2026	All other reportable incidents	47265	Dual-chamber implantable pacemaker, rate-responsive			E2403	No Clinical Signs, Symptoms or Conditions	A0908	Incorrect, Inadequate or Imprecise Result or Readings		
							F26	No Health Consequences or Impact	A0908	Incorrect, Inadequate or Imprecise Result or Readings		
18576	2026	All other reportable incidents	47265	Dual-chamber implantable pacemaker, rate-responsive			E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
18577	2026	All other reportable incidents	47265	Dual-chamber implantable pacemaker, rate-responsive			E020201	Anxiety	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
18578	2026	All other reportable incidents	47265	Dual-chamber implantable pacemaker, rate-responsive			E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
18579	2026	All other reportable incidents	47265	Dual-chamber implantable pacemaker, rate-responsive			E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
18580	2026	All other reportable incidents	47265	Dual-chamber implantable pacemaker, rate-responsive			E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
18581	2026	All other reportable incidents	47265	Dual-chamber implantable pacemaker, rate-responsive			E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
18582	2026	All other reportable incidents	47265	Dual-chamber implantable pacemaker, rate-responsive			E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
18583	2026	All other reportable incidents	47265	Dual-chamber implantable pacemaker, rate-responsive			E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
									A071102	Failure to Interrogate		
							F26	No Health Consequences or Impact	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
									A071102	Failure to Interrogate		
18584	2026	All other reportable incidents	47263	Cardiac resynchronization therapy implantable pacemaker	CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A0908	Incorrect, Inadequate or Imprecise Result or Readings		
							F26	No Health Consequences or Impact	A0908	Incorrect, Inadequate or Imprecise Result or Readings		
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A0908	Incorrect, Inadequate or Imprecise Result or Readings		
							F26	No Health Consequences or Impact	A0908	Incorrect, Inadequate or Imprecise Result or Readings		
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A0908	Incorrect, Inadequate or Imprecise Result or Readings		
							F26	No Health Consequences or Impact	A0908	Incorrect, Inadequate or Imprecise Result or Readings		
18585	2026		47265	Dual-chamber implantable pacemaker, rate-responsive	CT1685	Cardiac pacemakers	E02	Mental, Emotional and Behavioral Disorders	A071205	Pacemaker Found in Back-Up Mode		
							E020201	Anxiety	A071205	Pacemaker Found in Back-Up Mode		
							F1901	Additional Surgery	A071205	Pacemaker Found in Back-Up Mode		
					CT257	Implantable cardiac pacemakers	E02	Mental, Emotional and Behavioral Disorders	A071205	Pacemaker Found in Back-Up Mode		
							E020201	Anxiety	A071205	Pacemaker Found in Back-Up Mode		
							F1901	Additional Surgery	A071205	Pacemaker Found in Back-Up Mode		
					CT2866	Implantable cardiac electrical stimulators	E02	Mental, Emotional and Behavioral Disorders	A071205	Pacemaker Found in Back-Up Mode		
							E020201	Anxiety	A071205	Pacemaker Found in Back-Up Mode		
							F1901	Additional Surgery	A071205	Pacemaker Found in Back-Up Mode		

18586	2026	All other reportable incidents	35852	Single-chamber implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E020201	Anxiety	A071205	Pacemaker Found in Back-Up Mode	22/05/2026	22/05/2026
									A0908	Incorrect, Inadequate or Imprecise Result or Readings	22/05/2026	22/05/2026
									A0909	Unexpected Therapeutic Results	22/05/2026	22/05/2026
									A160503	Inappropriate or Unexpected Reset	22/05/2026	22/05/2026
							E2403	No Clinical Signs, Symptoms or Conditions	A071205	Pacemaker Found in Back-Up Mode	22/05/2026	22/05/2026
									A0908	Incorrect, Inadequate or Imprecise Result or Readings	22/05/2026	22/05/2026
									A0909	Unexpected Therapeutic Results	22/05/2026	22/05/2026
									A160503	Inappropriate or Unexpected Reset	22/05/2026	22/05/2026
							F01	Change in Therapeutic Response	A071205	Pacemaker Found in Back-Up Mode	22/05/2026	22/05/2026
									A0908	Incorrect, Inadequate or Imprecise Result or Readings	22/05/2026	22/05/2026
									A0909	Unexpected Therapeutic Results	22/05/2026	22/05/2026
									A160503	Inappropriate or Unexpected Reset	22/05/2026	22/05/2026
							F08	Hospitalization or Prolonged Hospitalization	A071205	Pacemaker Found in Back-Up Mode	22/05/2026	22/05/2026
									A0908	Incorrect, Inadequate or Imprecise Result or Readings	22/05/2026	22/05/2026
									A0909	Unexpected Therapeutic Results	22/05/2026	22/05/2026
									A160503	Inappropriate or Unexpected Reset	22/05/2026	22/05/2026
							F1901	Additional Surgery	A071205	Pacemaker Found in Back-Up Mode	22/05/2026	22/05/2026
									A0908	Incorrect, Inadequate or Imprecise Result or Readings	22/05/2026	22/05/2026
									A0909	Unexpected Therapeutic Results	22/05/2026	22/05/2026
									A160503	Inappropriate or Unexpected Reset	22/05/2026	22/05/2026
							F1905	Device Revision or Replacement	A071205	Pacemaker Found in Back-Up Mode	22/05/2026	22/05/2026
									A0908	Incorrect, Inadequate or Imprecise Result or Readings	22/05/2026	22/05/2026
									A0909	Unexpected Therapeutic Results	22/05/2026	22/05/2026
									A160503	Inappropriate or Unexpected Reset	22/05/2026	22/05/2026
18587	2026	All other reportable incidents	47265	Dual-chamber implantable pacemaker, rate-responsive			E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
18588	2026	All other reportable incidents	47265	Dual-chamber implantable pacemaker, rate-responsive	CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
18589	2026	All other reportable incidents	47265	Dual-chamber implantable pacemaker, rate-responsive			E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
18590	2026	All other reportable incidents	47265	Dual-chamber implantable pacemaker, rate-responsive	CT1685	Cardiac pacemakers	E020201	Anxiety	A070504	Premature Discharge of Battery	24/06/2026	24/06/2026
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	24/06/2026	24/06/2026
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	24/06/2026	24/06/2026
					CT257	Implantable cardiac pacemakers	E020201	Anxiety	A070504	Premature Discharge of Battery	24/06/2026	24/06/2026
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	24/06/2026	24/06/2026
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	24/06/2026	24/06/2026
					CT2866	Implantable cardiac electrical stimulators	E020201	Anxiety	A070504	Premature Discharge of Battery	24/06/2026	24/06/2026
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	24/06/2026	24/06/2026
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	24/06/2026	24/06/2026
18591	2026	All other reportable incidents	47265	Dual-chamber implantable pacemaker, rate-responsive			E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
18592	2026	All other reportable incidents	47265	Dual-chamber implantable pacemaker, rate-responsive			E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
18593	2026	All other reportable incidents	47265	Dual-chamber implantable pacemaker, rate-responsive			E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
18594	2026	All other reportable incidents	47265	Dual-chamber implantable pacemaker, rate-responsive			E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
18595	2026	All other reportable incidents	47265	Dual-chamber implantable pacemaker, rate-responsive	CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
18596	2026	All other reportable incidents	47265	Dual-chamber implantable pacemaker, rate-responsive			E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
18597	2026	All other reportable incidents	47267	Single-chamber implantable pacemaker, rate-responsive			E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
18598	2026	All other reportable incidents	47265	Dual-chamber implantable pacemaker, rate-responsive			E2403	No Clinical Signs, Symptoms or Conditions	A0908	Incorrect, Inadequate or Imprecise Result or Readings	21/05/2026	21/05/2026
							F08	Hospitalization or Prolonged Hospitalization	A0908	Incorrect, Inadequate or Imprecise Result or Readings	21/05/2026	21/05/2026
							F1905	Device Revision or Replacement	A0908	Incorrect, Inadequate or Imprecise Result or Readings	21/05/2026	21/05/2026
18599	2026	All other reportable incidents	47265	Dual-chamber implantable pacemaker, rate-responsive			E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
18600	2026	All other reportable incidents	47265	Dual-chamber implantable pacemaker, rate-responsive	CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
18601	2026	All other reportable incidents	47265	Dual-chamber implantable pacemaker, rate-responsive			E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
18602	2026	All other reportable incidents	47265	Dual-chamber implantable pacemaker, rate-responsive			E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
18603	2026	All other reportable incidents	47265	Dual-chamber implantable pacemaker, rate-responsive			E2403	No Clinical Signs, Symptoms or Conditions	A05	Mechanical Problem	29/05/2026	29/05/2026
							F26	No Health Consequences or Impact	A05	Mechanical Problem	29/05/2026	29/05/2026
18604	2026	Unanticipated serious deterioration in state of health	47267	Single-chamber implantable pacemaker, rate-responsive			E060101	Asystole	A070101	Failure to Capture	29/05/2026	29/05/2026
									A0712	Pacing Problem	29/05/2026	29/05/2026
							F1203	Life Threatening Illness or Injury	A070101	Failure to Capture	29/05/2026	29/05/2026

									A0712	Pacing Problem	29/05/2026	29/05/2026
18605	2026	All other reportable incidents	36106	Implantable pacemaker analyser, intraoperative	CT1684	Cardiac pacemakers and associated devices	E011902	Presyncope	A070504	Premature Discharge of Battery		
									A070909	Oversensing		
									A0722	Impedance Problem		
									A072202	Low Impedance		
									A160602	Premature End-of-Life Indicator		
							E060104	Bradycardia	A070504	Premature Discharge of Battery		
									A070909	Oversensing		
									A0722	Impedance Problem		
									A072202	Low Impedance		
									A160602	Premature End-of-Life Indicator		
							E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
									A070909	Oversensing		
									A0722	Impedance Problem		
									A072202	Low Impedance		
									A160602	Premature End-of-Life Indicator		
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery		
									A070909	Oversensing		
									A0722	Impedance Problem		
									A072202	Low Impedance		
									A160602	Premature End-of-Life Indicator		
							F1205	Temporary Impairment	A070504	Premature Discharge of Battery		
									A070909	Oversensing		
									A0722	Impedance Problem		
									A072202	Low Impedance		
									A160602	Premature End-of-Life Indicator		
							F19	Surgical Intervention	A070504	Premature Discharge of Battery		
									A070909	Oversensing		
									A0722	Impedance Problem		
									A072202	Low Impedance		
									A160602	Premature End-of-Life Indicator		
							F1901	Additional Surgery	A070504	Premature Discharge of Battery		
									A070909	Oversensing		
									A0722	Impedance Problem		
									A072202	Low Impedance		
									A160602	Premature End-of-Life Indicator		
							F23	Unexpected Medical Intervention	A070504	Premature Discharge of Battery		
									A070909	Oversensing		
									A0722	Impedance Problem		
									A072202	Low Impedance		
									A160602	Premature End-of-Life Indicator		
18606	2026	All other reportable incidents	47265	Dual-chamber implantable pacemaker, rate-responsive	CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A160105	False Alarm		
							F26	No Health Consequences or Impact	A160105	False Alarm		
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A160105	False Alarm		
							F26	No Health Consequences or Impact	A160105	False Alarm		
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A160105	False Alarm		
							F26	No Health Consequences or Impact	A160105	False Alarm		
18607	2026	All other reportable incidents	60789	Intracardiac pacemaker			E2403	No Clinical Signs, Symptoms or Conditions	A070102	High Capture Threshold		
									A070504	Premature Discharge of Battery		
							F12	Serious Injury/ Illness/ Impairment	A070102	High Capture Threshold		
									A070504	Premature Discharge of Battery		
							F2307	Reprogramming of Device	A070102	High Capture Threshold		
									A070504	Premature Discharge of Battery		
18608	2026	All other reportable incidents	47265	Dual-chamber implantable pacemaker, rate-responsive			E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
18609	2026	All other reportable incidents	47265	Dual-chamber implantable pacemaker, rate-responsive			E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
18610	2026	All other reportable incidents	60789	Intracardiac pacemaker			E0119	Loss of Consciousness	A070101	Failure to Capture	01/06/2026	30/06/2026
									A070102	High Capture Threshold	01/06/2026	30/06/2026
									A0712	Pacing Problem	01/06/2026	30/06/2026
							E0618	Non Specific EKG/ECG Changes	A070101	Failure to Capture	01/06/2026	30/06/2026
									A070102	High Capture Threshold	01/06/2026	30/06/2026
									A0712	Pacing Problem	01/06/2026	30/06/2026
							F08	Hospitalization or Prolonged Hospitalization	A070101	Failure to Capture	01/06/2026	30/06/2026
									A070102	High Capture Threshold	01/06/2026	30/06/2026
									A0712	Pacing Problem	01/06/2026	30/06/2026
							F12	Serious Injury/ Illness/ Impairment	A070101	Failure to Capture	01/06/2026	30/06/2026
									A070102	High Capture Threshold	01/06/2026	30/06/2026
									A0712	Pacing Problem	01/06/2026	30/06/2026
							F1203	Life Threatening Illness or Injury	A070101	Failure to Capture	01/06/2026	30/06/2026
									A070102	High Capture Threshold	01/06/2026	30/06/2026
									A0712	Pacing Problem	01/06/2026	30/06/2026
							F1903	Device Explantation	A070101	Failure to Capture	01/06/2026	30/06/2026
									A070102	High Capture Threshold	01/06/2026	30/06/2026
									A0712	Pacing Problem	01/06/2026	30/06/2026
							F2307	Reprogramming of Device	A070101	Failure to Capture	01/06/2026	30/06/2026
									A070102	High Capture Threshold	01/06/2026	30/06/2026
									A0712	Pacing Problem	01/06/2026	30/06/2026
18611	2026	All other reportable incidents	60789	Intracardiac pacemaker			E2403	No Clinical Signs, Symptoms or Conditions	A070101	Failure to Capture	03/06/2026	03/06/2026
									A070901	Decreased Sensitivity	03/06/2026	03/06/2026
							F26	No Health Consequences or Impact	A070101	Failure to Capture	03/06/2026	03/06/2026
									A070901	Decreased Sensitivity	03/06/2026	03/06/2026
18612	2026	All other reportable incidents	47265	Dual-chamber implantable pacemaker, rate-responsive			E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
18613	2026	All other reportable incidents	47265	Dual-chamber implantable pacemaker, rate-responsive			E0130	Sleep Dysfunction	A26	Insufficient Device Problem Information		
							F11	Non-Serious Injury/ Illness/ Impairment	A26	Insufficient Device Problem Information		
18614	2026	All other reportable incidents	47265	Dual-chamber implantable pacemaker, rate-responsive			E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		

18615	2026	All other reportable incidents	47265	Dual-chamber implantable pacemaker, rate-responsive			E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
18616	2026	All other reportable incidents	47265	Dual-chamber implantable pacemaker, rate-responsive	CT1685	Cardiac pacemakers	E0112	Dizziness				
							E0601	Arrhythmia				
							F1908	Prolonged Surgery				
					CT257	Implantable cardiac pacemakers	E0112	Dizziness				
							E0601	Arrhythmia				
							F1908	Prolonged Surgery				
					CT2866	Implantable cardiac electrical stimulators	E0112	Dizziness				
							E0601	Arrhythmia				
							F1908	Prolonged Surgery				
18617	2026	All other reportable incidents	47265	Dual-chamber implantable pacemaker, rate-responsive			E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
18618	2026	All other reportable incidents	47265	Dual-chamber implantable pacemaker, rate-responsive			E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
18619	2026	All other reportable incidents	47265	Dual-chamber implantable pacemaker, rate-responsive			E2403	No Clinical Signs, Symptoms or Conditions	A05	Mechanical Problem	05/06/2026	05/06/2026
							F26	No Health Consequences or Impact	A05	Mechanical Problem	05/06/2026	05/06/2026
18620	2026	All other reportable incidents	47265	Dual-chamber implantable pacemaker, rate-responsive	CT1685	Cardiac pacemakers	E011902	Presyncope	A0203	Defective Device		
									A071102	Failure to Interrogate		
									A0712	Pacing Problem		
									A071204	No Pacing		
							E2403	No Clinical Signs, Symptoms or Conditions	A0203	Defective Device		
									A071102	Failure to Interrogate		
									A0712	Pacing Problem		
									A071204	No Pacing		
							F08	Hospitalization or Prolonged Hospitalization	A0203	Defective Device		
									A071102	Failure to Interrogate		
									A0712	Pacing Problem		
									A071204	No Pacing		
							F11	Non-Serious Injury/ Illness/ Impairment	A0203	Defective Device		
									A071102	Failure to Interrogate		
									A0712	Pacing Problem		
									A071204	No Pacing		
							F1901	Additional Surgery	A0203	Defective Device		
									A071102	Failure to Interrogate		
									A0712	Pacing Problem		
									A071204	No Pacing		
							F1905	Device Revision or Replacement	A0203	Defective Device		
									A071102	Failure to Interrogate		
									A0712	Pacing Problem		
									A071204	No Pacing		
							F23	Unexpected Medical Intervention	A0203	Defective Device		
									A071102	Failure to Interrogate		
									A0712	Pacing Problem		
									A071204	No Pacing		
					CT257	Implantable cardiac pacemakers	E011902	Presyncope	A0203	Defective Device		
									A071102	Failure to Interrogate		
									A0712	Pacing Problem		
									A071204	No Pacing		
							E2403	No Clinical Signs, Symptoms or Conditions	A0203	Defective Device		
									A071102	Failure to Interrogate		
									A0712	Pacing Problem		
									A071204	No Pacing		
							F08	Hospitalization or Prolonged Hospitalization	A0203	Defective Device		
									A071102	Failure to Interrogate		
									A0712	Pacing Problem		
									A071204	No Pacing		
							F11	Non-Serious Injury/ Illness/ Impairment	A0203	Defective Device		
									A071102	Failure to Interrogate		
									A0712	Pacing Problem		
									A071204	No Pacing		
							F1901	Additional Surgery	A0203	Defective Device		
									A071102	Failure to Interrogate		
									A0712	Pacing Problem		
									A071204	No Pacing		
							F1905	Device Revision or Replacement	A0203	Defective Device		
									A071102	Failure to Interrogate		
									A0712	Pacing Problem		
									A071204	No Pacing		
							F23	Unexpected Medical Intervention	A0203	Defective Device		
									A071102	Failure to Interrogate		
									A0712	Pacing Problem		
									A071204	No Pacing		
					CT2866	Implantable cardiac electrical stimulators	E011902	Presyncope	A0203	Defective Device		
									A071102	Failure to Interrogate		
									A0712	Pacing Problem		
									A071204	No Pacing		
							E2403	No Clinical Signs, Symptoms or Conditions	A0203	Defective Device		
									A071102	Failure to Interrogate		
									A0712	Pacing Problem		
									A071204	No Pacing		
							F08	Hospitalization or Prolonged Hospitalization	A0203	Defective Device		
									A071102	Failure to Interrogate		
									A0712	Pacing Problem		
									A071204	No Pacing		
							F11	Non-Serious Injury/ Illness/ Impairment	A0203	Defective Device		
									A071102	Failure to Interrogate		
									A0712	Pacing Problem		

									A071204	No Pacing		
						F1901	Additional Surgery		A0203	Defective Device		
									A071102	Failure to Interrogate		
									A0712	Pacing Problem		
									A071204	No Pacing		
						F1905	Device Revision or Replacement		A0203	Defective Device		
									A071102	Failure to Interrogate		
									A0712	Pacing Problem		
									A071204	No Pacing		
						F23	Unexpected Medical Intervention		A0203	Defective Device		
									A071102	Failure to Interrogate		
									A0712	Pacing Problem		
									A071204	No Pacing		
18621	2026		47263	Cardiac resynchronization therapy implantable pacemaker	CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A0203	Defective Device		
									A070501	Battery Problem: High Impedance		
									A0722	Impedance Problem		
							F1905	Device Revision or Replacement	A0203	Defective Device		
									A070501	Battery Problem: High Impedance		
									A0722	Impedance Problem		
							F26	No Health Consequences or Impact	A0203	Defective Device		
									A070501	Battery Problem: High Impedance		
									A0722	Impedance Problem		
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A0203	Defective Device		
									A070501	Battery Problem: High Impedance		
									A0722	Impedance Problem		
							F1905	Device Revision or Replacement	A0203	Defective Device		
									A070501	Battery Problem: High Impedance		
									A0722	Impedance Problem		
							F26	No Health Consequences or Impact	A0203	Defective Device		
									A070501	Battery Problem: High Impedance		
									A0722	Impedance Problem		
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A0203	Defective Device		
									A070501	Battery Problem: High Impedance		
									A0722	Impedance Problem		
							F1905	Device Revision or Replacement	A0203	Defective Device		
									A070501	Battery Problem: High Impedance		
									A0722	Impedance Problem		
							F26	No Health Consequences or Impact	A0203	Defective Device		
									A070501	Battery Problem: High Impedance		
									A0722	Impedance Problem		
18622	2026	All other reportable incidents	47265	Dual-chamber implantable pacemaker, rate-responsive			E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
18623	2026	All other reportable incidents	47267	Single-chamber implantable pacemaker, rate-responsive			E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
18624	2026	All other reportable incidents	47265	Dual-chamber implantable pacemaker, rate-responsive			E2403	No Clinical Signs, Symptoms or Conditions	A070909	Oversensing		
									A070910	Undersensing		
									A0712	Pacing Problem		
									A090801	Signal Artifact/Noise		
							F15	Recognized Device or Procedural Complication	A070909	Oversensing		
									A070910	Undersensing		
									A0712	Pacing Problem		
									A090801	Signal Artifact/Noise		
							F1009	Inappropriate Treatment	A070909	Oversensing		
									A070910	Undersensing		
									A0712	Pacing Problem		
									A090801	Signal Artifact/Noise		
18625	2026	All other reportable incidents	47265	Dual-chamber implantable pacemaker, rate-responsive			E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
18626	2026	All other reportable incidents	47265	Dual-chamber implantable pacemaker, rate-responsive			E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
18627	2026	All other reportable incidents	35852	Single-chamber implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A0709	Device Sensing Problem		
									A070909	Oversensing		
									A090801	Signal Artifact/Noise		
							F2203	Imaging Required	A0709	Device Sensing Problem		
									A070909	Oversensing		
									A090801	Signal Artifact/Noise		
							F2307	Reprogramming of Device	A0709	Device Sensing Problem		
									A070909	Oversensing		
									A090801	Signal Artifact/Noise		
18628	2026	All other reportable incidents	60956	Pacing/defibrillation lead tunneller	CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A040101	Fracture		
									A0709	Device Sensing Problem		
									A070909	Oversensing		
									A090801	Signal Artifact/Noise		
							F2203	Imaging Required	A040101	Fracture		
									A0709	Device Sensing Problem		
									A070909	Oversensing		
									A090801	Signal Artifact/Noise		
							F2307	Reprogramming of Device	A040101	Fracture		
									A0709	Device Sensing Problem		
									A070909	Oversensing		
									A090801	Signal Artifact/Noise		
18629	2026		17882	Physiologic-monitoring defibrillation system	CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A070802	Intermittent Loss of Power		
									A090203	Erratic or Intermittent Display		
							F25	Unanticipated Adverse Device Effect	A070802	Intermittent Loss of Power		
									A090203	Erratic or Intermittent Display		
18630	2026	All other reportable incidents	47265	Dual-chamber implantable pacemaker, rate-responsive	CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		

							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
18631	2026	All other reportable incidents	47265	Dual-chamber implantable pacemaker, rate-responsive			E2403	No Clinical Signs, Symptoms or Conditions	A070101	Failure to Capture	22/05/2026	22/05/2026
							F1901	Additional Surgery	A070101	Failure to Capture	22/05/2026	22/05/2026
18632	2026	All other reportable incidents	47265	Dual-chamber implantable pacemaker, rate-responsive			E2403	No Clinical Signs, Symptoms or Conditions	A0909	Unexpected Therapeutic Results	30/03/2026	30/03/2026
							F08	Hospitalization or Prolonged Hospitalization	A0909	Unexpected Therapeutic Results	30/03/2026	30/03/2026
							F1905	Device Revision or Replacement	A0909	Unexpected Therapeutic Results	30/03/2026	30/03/2026
18633	2026	All other reportable incidents	47265	Dual-chamber implantable pacemaker, rate-responsive			E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
18634	2026	All other reportable incidents	47265	Dual-chamber implantable pacemaker, rate-responsive			E2403	No Clinical Signs, Symptoms or Conditions	A0711	Interrogation Problem		
									A23	Use of Device Problem		
							F26	No Health Consequences or Impact	A0711	Interrogation Problem		
									A23	Use of Device Problem		
18635	2026	All other reportable incidents	35852	Single-chamber implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E020201	Anxiety	A070504	Premature Discharge of Battery		
									A27	Appropriate Device Problem Term/Code not Available		
							F11	Non-Serious Injury/ Illness/ Impairment	A070504	Premature Discharge of Battery		
									A27	Appropriate Device Problem Term/Code not Available		
							F1901	Additional Surgery	A070504	Premature Discharge of Battery		
									A27	Appropriate Device Problem Term/Code not Available		
18636	2026	All other reportable incidents	47265	Dual-chamber implantable pacemaker, rate-responsive	CT1685	Cardiac pacemakers	E020201	Anxiety	A070504	Premature Discharge of Battery		
									A27	Appropriate Device Problem Term/Code not Available		
							F1901	Additional Surgery	A070504	Premature Discharge of Battery		
									A27	Appropriate Device Problem Term/Code not Available		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
									A27	Appropriate Device Problem Term/Code not Available		
					CT257	Implantable cardiac pacemakers	E020201	Anxiety	A070504	Premature Discharge of Battery		
									A27	Appropriate Device Problem Term/Code not Available		
							F1901	Additional Surgery	A070504	Premature Discharge of Battery		
									A27	Appropriate Device Problem Term/Code not Available		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
									A27	Appropriate Device Problem Term/Code not Available		
					CT2866	Implantable cardiac electrical stimulators	E020201	Anxiety	A070504	Premature Discharge of Battery		
									A27	Appropriate Device Problem Term/Code not Available		
							F1901	Additional Surgery	A070504	Premature Discharge of Battery		
									A27	Appropriate Device Problem Term/Code not Available		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
									A27	Appropriate Device Problem Term/Code not Available		
18637	2026	All other reportable incidents	47265	Dual-chamber implantable pacemaker, rate-responsive			E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
18638	2026		44863	Temporary cardiac pacing/stimulation connection cable, single-use	CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A0203	Defective Device		
									A1406	Inflation Problem		
									A27	Appropriate Device Problem Term/Code not Available		
							F26	No Health Consequences or Impact	A0203	Defective Device		
									A1406	Inflation Problem		
									A27	Appropriate Device Problem Term/Code not Available		
18639	2026	All other reportable incidents	47263	Cardiac resynchronization therapy implantable pacemaker	CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A070102	High Capture Threshold		
									A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070102	High Capture Threshold		
									A070504	Premature Discharge of Battery		
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A070102	High Capture Threshold		
									A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070102	High Capture Threshold		
									A070504	Premature Discharge of Battery		
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A070102	High Capture Threshold		
									A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070102	High Capture Threshold		
									A070504	Premature Discharge of Battery		
18640	2026	All other reportable incidents	47265	Dual-chamber implantable pacemaker, rate-responsive			E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
18641	2026	All other reportable incidents	47265	Dual-chamber implantable pacemaker, rate-responsive			E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
18642	2026	All other reportable incidents	47265	Dual-chamber implantable pacemaker, rate-responsive			E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
18643	2026	All other reportable incidents	47265	Dual-chamber implantable pacemaker, rate-responsive			E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
18644	2026	All other reportable incidents	47265	Dual-chamber implantable pacemaker, rate-responsive	CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A070909	Oversensing		
									A070910	Undersensing		
									A1301	Failure to Read Input Signal		
							F26	No Health Consequences or Impact	A070909	Oversensing		
									A070910	Undersensing		
									A1301	Failure to Read Input Signal		
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A070909	Oversensing		
									A070910	Undersensing		
									A1301	Failure to Read Input Signal		
							F26	No Health Consequences or Impact	A070909	Oversensing		
									A070910	Undersensing		

									A1301	Failure to Read Input Signal		
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A070909	Oversensing		
									A070910	Undersensing		
									A1301	Failure to Read Input Signal		
							F26	No Health Consequences or Impact	A070909	Oversensing		
									A070910	Undersensing		
									A1301	Failure to Read Input Signal		
18645	2026	Unanticipated serious deterioration in state of health	60789	Intracardiac pacemaker			E0619	Pericardial Effusion	A070102	High Capture Threshold		
									A070901	Decreased Sensitivity		
									A1502	Positioning Problem		
							F08	Hospitalization or Prolonged Hospitalization	A070102	High Capture Threshold		
									A070901	Decreased Sensitivity		
									A1502	Positioning Problem		
							F1203	Life Threatening Illness or Injury	A070102	High Capture Threshold		
									A070901	Decreased Sensitivity		
									A1502	Positioning Problem		
							F23	Unexpected Medical Intervention	A070102	High Capture Threshold		
									A070901	Decreased Sensitivity		
									A1502	Positioning Problem		
18646	2026	All other reportable incidents	47265	Dual-chamber implantable pacemaker, rate-responsive			E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
18647	2026	All other reportable incidents	47265	Dual-chamber implantable pacemaker, rate-responsive			E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
18648	2026	All other reportable incidents	47205	Cardiac pulse generator programmer	CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
18649	2026	All other reportable incidents	47265	Dual-chamber implantable pacemaker, rate-responsive			E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
18650	2026	All other reportable incidents	47265	Dual-chamber implantable pacemaker, rate-responsive			E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
18651	2026	All other reportable incidents	47265	Dual-chamber implantable pacemaker, rate-responsive			E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
18652	2026	All other reportable incidents	47265	Dual-chamber implantable pacemaker, rate-responsive			E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
18653	2026	All other reportable incidents	47265	Dual-chamber implantable pacemaker, rate-responsive			E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
18654	2026	All other reportable incidents	47265	Dual-chamber implantable pacemaker, rate-responsive			E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
18655	2026	All other reportable incidents	60789	Intracardiac pacemaker			E2403	No Clinical Signs, Symptoms or Conditions	A070102	High Capture Threshold		
									A0712	Pacing Problem		
							F2307	Reprogramming of Device	A070102	High Capture Threshold		
									A0712	Pacing Problem		
18656	2026	All other reportable incidents	47267	Single-chamber implantable pacemaker, rate-responsive			E2403	No Clinical Signs, Symptoms or Conditions	A0205	Packaging Problem		
							F26	No Health Consequences or Impact	A0205	Packaging Problem		
18657	2026	All other reportable incidents	47265	Dual-chamber implantable pacemaker, rate-responsive			E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
18658	2026	All other reportable incidents	47265	Dual-chamber implantable pacemaker, rate-responsive			E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
18659	2026	All other reportable incidents	47265	Dual-chamber implantable pacemaker, rate-responsive			E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
18660	2026	All other reportable incidents	47265	Dual-chamber implantable pacemaker, rate-responsive			E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
18661	2026	All other reportable incidents	47265	Dual-chamber implantable pacemaker, rate-responsive			E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
18662	2026	All other reportable incidents	47265	Dual-chamber implantable pacemaker, rate-responsive			E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
18663	2026	All other reportable incidents	47265	Dual-chamber implantable pacemaker, rate-responsive			E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
18664	2026	All other reportable incidents	47265	Dual-chamber implantable pacemaker, rate-responsive			E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
18665	2026	All other reportable incidents	47265	Dual-chamber implantable pacemaker, rate-responsive			E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
									A072201	High Impedance		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
									A072201	High Impedance		
18666	2026	All other reportable incidents	60789	Intracardiac pacemaker			E2403	No Clinical Signs, Symptoms or Conditions	A070101	Failure to Capture	05/06/2026	05/06/2026
							F1901	Additional Surgery	A070101	Failure to Capture	05/06/2026	05/06/2026
18667	2026	All other reportable incidents	60789	Intracardiac pacemaker			E0601	Arrhythmia	A070101	Failure to Capture	14/06/2026	14/06/2026
							E060106	Heart Block	A070101	Failure to Capture	14/06/2026	14/06/2026
							F08	Hospitalization or Prolonged Hospitalization	A070101	Failure to Capture	14/06/2026	14/06/2026
							F1905	Device Revision or Replacement	A070101	Failure to Capture	14/06/2026	14/06/2026
18668	2026		47265	Dual-chamber implantable pacemaker, rate-responsive	CT1685	Cardiac pacemakers	E011903	Syncope/Fainting	A170101	Accessory Incompatible		
							F1203	Life Threatening Illness or Injury	A170101	Accessory Incompatible		
					CT257	Implantable cardiac pacemakers	E011903	Syncope/Fainting	A170101	Accessory Incompatible		
							F1203	Life Threatening Illness or Injury	A170101	Accessory Incompatible		
					CT2866	Implantable cardiac electrical stimulators	E011903	Syncope/Fainting	A170101	Accessory Incompatible		
							F1203	Life Threatening Illness or Injury	A170101	Accessory Incompatible		
18669	2026	All other reportable incidents	47265	Dual-chamber implantable pacemaker, rate-responsive			E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
18670	2026	All other reportable incidents	47265	Dual-chamber implantable pacemaker, rate-responsive			E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
18671	2026	All other reportable incidents	47265	Dual-chamber implantable pacemaker, rate-responsive			E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
									A0711	Interrogation Problem		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
									A0711	Interrogation Problem		
18672	2026	All other reportable incidents	47265	Dual-chamber implantable pacemaker, rate-responsive			E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
18673	2026	All other reportable incidents	47265	Dual-chamber implantable pacemaker, rate-			E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		

				responsive			F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
18674	2026	All other reportable incidents	47265	Dual-chamber implantable pacemaker, rate-responsive			E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
18675	2026	All other reportable incidents	47265	Dual-chamber implantable pacemaker, rate-responsive	CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A070909	Oversensing		
									A090801	Signal Artifact/Noise		
							F26	No Health Consequences or Impact	A070909	Oversensing		
									A090801	Signal Artifact/Noise		
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A070909	Oversensing		
									A090801	Signal Artifact/Noise		
							F26	No Health Consequences or Impact	A070909	Oversensing		
									A090801	Signal Artifact/Noise		
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A070909	Oversensing		
									A090801	Signal Artifact/Noise		
							F26	No Health Consequences or Impact	A070909	Oversensing		
									A090801	Signal Artifact/Noise		
18676	2026	All other reportable incidents	47265	Dual-chamber implantable pacemaker, rate-responsive			E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
18677	2026	All other reportable incidents	47267	Single-chamber implantable pacemaker, rate-responsive			E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
18678	2026	All other reportable incidents	47265	Dual-chamber implantable pacemaker, rate-responsive			E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
18679	2026	All other reportable incidents	47265	Dual-chamber implantable pacemaker, rate-responsive			E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
18680	2026	All other reportable incidents	47265	Dual-chamber implantable pacemaker, rate-responsive			E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
18681	2026	All other reportable incidents	47265	Dual-chamber implantable pacemaker, rate-responsive			E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
18682	2026		35852	Single-chamber implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F1901	Additional Surgery	A070504	Premature Discharge of Battery		
18683	2026		35852	Single-chamber implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A04	Material Integrity Problem		
									A0705	Battery Problem		
							F23	Unexpected Medical Intervention	A04	Material Integrity Problem		
									A0705	Battery Problem		
18684	2026		47265	Dual-chamber implantable pacemaker, rate-responsive	CT1685	Cardiac pacemakers	E020201	Anxiety	A1102	Application Program Problem		
									A27	Appropriate Device Problem Term/Code not Available		
							F1901	Additional Surgery	A1102	Application Program Problem		
									A27	Appropriate Device Problem Term/Code not Available		
					CT257	Implantable cardiac pacemakers	E020201	Anxiety	A1102	Application Program Problem		
									A27	Appropriate Device Problem Term/Code not Available		
							F1901	Additional Surgery	A1102	Application Program Problem		
									A27	Appropriate Device Problem Term/Code not Available		
					CT2750	Implantable cardiac electrical stimulators and associated devices	E020201	Anxiety	A1102	Application Program Problem		
									A27	Appropriate Device Problem Term/Code not Available		
							F1901	Additional Surgery	A1102	Application Program Problem		
									A27	Appropriate Device Problem Term/Code not Available		
					CT2866	Implantable cardiac electrical stimulators	E020201	Anxiety	A1102	Application Program Problem		
									A27	Appropriate Device Problem Term/Code not Available		
							F1901	Additional Surgery	A1102	Application Program Problem		
									A27	Appropriate Device Problem Term/Code not Available		
18685	2026		47265	Dual-chamber implantable pacemaker, rate-responsive	CT1685	Cardiac pacemakers	E020201	Anxiety	A071205	Pacemaker Found in Back-Up Mode		
									A27	Appropriate Device Problem Term/Code not Available		
							F1901	Additional Surgery	A071205	Pacemaker Found in Back-Up Mode		
									A27	Appropriate Device Problem Term/Code not Available		
					CT257	Implantable cardiac pacemakers	E020201	Anxiety	A071205	Pacemaker Found in Back-Up Mode		
									A27	Appropriate Device Problem Term/Code not Available		
							F1901	Additional Surgery	A071205	Pacemaker Found in Back-Up Mode		
									A27	Appropriate Device Problem Term/Code not Available		
					CT2750	Implantable cardiac electrical stimulators and associated devices	E020201	Anxiety	A071205	Pacemaker Found in Back-Up Mode		
									A27	Appropriate Device Problem Term/Code not Available		
							F1901	Additional Surgery	A071205	Pacemaker Found in Back-Up Mode		
									A27	Appropriate Device Problem Term/Code not Available		
					CT2866	Implantable cardiac electrical stimulators	E020201	Anxiety	A071205	Pacemaker Found in Back-Up Mode		
									A27	Appropriate Device Problem Term/Code not Available		
							F1901	Additional Surgery	A071205	Pacemaker Found in Back-Up Mode		
									A27	Appropriate Device Problem Term/Code not Available		
18686	2026		37265	Dual-chamber implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E2401	Insufficient Information	A07	Electrical /Electronic Property Problem		
							F23	Unexpected Medical Intervention	A07	Electrical /Electronic Property Problem		
18687	2026	All other reportable incidents	47265	Dual-chamber implantable pacemaker, rate-responsive			E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	15/06/2026	15/06/2026
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	15/06/2026	15/06/2026
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	15/06/2026	15/06/2026
18688	2026	All other reportable incidents	47265	Dual-chamber implantable pacemaker, rate-responsive			E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	15/06/2026	15/06/2026
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	15/06/2026	15/06/2026
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	15/06/2026	15/06/2026
18689	2026	All other reportable incidents	47265	Dual-chamber implantable pacemaker, rate-responsive			E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
18690	2026	All other reportable incidents	47265	Dual-chamber implantable pacemaker, rate-			E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		

				responsive			F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
18691	2026	All other reportable incidents	47265	Dual-chamber implantable pacemaker, rate-responsive			E2403	No Clinical Signs, Symptoms or Conditions	A070909	Oversensing		
									A070910	Undersensing		
							F26	No Health Consequences or Impact	A070909	Oversensing		
									A070910	Undersensing		
18692	2026	All other reportable incidents	47267	Single-chamber implantable pacemaker, rate-responsive			E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	15/06/2026	15/06/2026
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	15/06/2026	15/06/2026
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	15/06/2026	15/06/2026
18693	2026	All other reportable incidents	47265	Dual-chamber implantable pacemaker, rate-responsive	CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A070910	Undersensing		
									A090801	Signal Artifact/Noise		
									A1107	Data Problem		
							F26	No Health Consequences or Impact	A070910	Undersensing		
									A090801	Signal Artifact/Noise		
									A1107	Data Problem		
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A070910	Undersensing		
									A090801	Signal Artifact/Noise		
									A1107	Data Problem		
							F26	No Health Consequences or Impact	A070910	Undersensing		
									A090801	Signal Artifact/Noise		
									A1107	Data Problem		
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A070910	Undersensing		
									A090801	Signal Artifact/Noise		
									A1107	Data Problem		
							F26	No Health Consequences or Impact	A070910	Undersensing		
									A090801	Signal Artifact/Noise		
									A1107	Data Problem		
18694	2026	All other reportable incidents	47265	Dual-chamber implantable pacemaker, rate-responsive			E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
18695	2026	All other reportable incidents	37265	Dual-chamber implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E210401	Shock from Patient Lead(s)	A070909	Oversensing		
									A071302	Inappropriate/Inadequate Shock/Stimulation		
							F1009	Inappropriate Treatment	A070909	Oversensing		
									A071302	Inappropriate/Inadequate Shock/Stimulation		
18696	2026	All other reportable incidents	47265	Dual-chamber implantable pacemaker, rate-responsive			E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
18697	2026	All other reportable incidents	47265	Dual-chamber implantable pacemaker, rate-responsive			E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
18698	2026		35822	External pacemaker, transcutaneous pacing	CT1685	Cardiac pacemakers	E060101	Asystole	A0705	Battery Problem		
									A0712	Pacing Problem		
							F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A0705	Battery Problem		
									A0712	Pacing Problem		
18699	2026	All other reportable incidents	47265	Dual-chamber implantable pacemaker, rate-responsive			E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery	15/06/2026	15/06/2026
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	15/06/2026	15/06/2026
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	15/06/2026	15/06/2026
18700	2026	All other reportable incidents	47265	Dual-chamber implantable pacemaker, rate-responsive			E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
18701	2026	All other reportable incidents	47265	Dual-chamber implantable pacemaker, rate-responsive			E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
18702	2026	All other reportable incidents	47265	Dual-chamber implantable pacemaker, rate-responsive			E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
18703	2026	All other reportable incidents	47265	Dual-chamber implantable pacemaker, rate-responsive			E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
18704	2026	All other reportable incidents	47265	Dual-chamber implantable pacemaker, rate-responsive			E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
18705	2026	All other reportable incidents	47267	Single-chamber implantable pacemaker, rate-responsive			E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
18706	2026	All other reportable incidents	47265	Dual-chamber implantable pacemaker, rate-responsive	CT1685	Cardiac pacemakers	E020201	Anxiety	A0708	Power Problem		
									A071205	Pacemaker Found in Back-Up Mode		
									A0908	Incorrect, Inadequate or Imprecise Result or Readings		
									A13	Communication or Transmission Problem		
							E2403	No Clinical Signs, Symptoms or Conditions	A0708	Power Problem		
									A071205	Pacemaker Found in Back-Up Mode		
									A0908	Incorrect, Inadequate or Imprecise Result or Readings		
									A13	Communication or Transmission Problem		
							F01	Change in Therapeutic Response	A0708	Power Problem		
									A071205	Pacemaker Found in Back-Up Mode		
									A0908	Incorrect, Inadequate or Imprecise Result or Readings		
									A13	Communication or Transmission Problem		
							F0101	Therapeutic Response Decreased	A0708	Power Problem		
									A071205	Pacemaker Found in Back-Up Mode		
									A0908	Incorrect, Inadequate or Imprecise Result or Readings		
									A13	Communication or Transmission Problem		
							F08	Hospitalization or Prolonged Hospitalization	A0708	Power Problem		
									A071205	Pacemaker Found in Back-Up Mode		
									A0908	Incorrect, Inadequate or Imprecise Result or Readings		
									A13	Communication or Transmission Problem		
							F1901	Additional Surgery	A0708	Power Problem		
									A071205	Pacemaker Found in Back-Up Mode		
									A0908	Incorrect, Inadequate or Imprecise Result or Readings		
									A13	Communication or Transmission Problem		
					CT257	Implantable cardiac pacemakers	E020201	Anxiety	A0708	Power Problem		
									A071205	Pacemaker Found in Back-Up Mode		
									A0908	Incorrect, Inadequate or Imprecise Result or Readings		

									A13	Communication or Transmission Problem			
							E2403	No Clinical Signs, Symptoms or Conditions	A0708	Power Problem			
									A071205	Pacemaker Found in Back-Up Mode			
									A0908	Incorrect, Inadequate or Imprecise Result or Readings			
									A13	Communication or Transmission Problem			
							F01	Change in Therapeutic Response	A0708	Power Problem			
									A071205	Pacemaker Found in Back-Up Mode			
									A0908	Incorrect, Inadequate or Imprecise Result or Readings			
									A13	Communication or Transmission Problem			
							F0101	Therapeutic Response Decreased	A0708	Power Problem			
									A071205	Pacemaker Found in Back-Up Mode			
									A0908	Incorrect, Inadequate or Imprecise Result or Readings			
									A13	Communication or Transmission Problem			
							F08	Hospitalization or Prolonged Hospitalization	A0708	Power Problem			
									A071205	Pacemaker Found in Back-Up Mode			
									A0908	Incorrect, Inadequate or Imprecise Result or Readings			
									A13	Communication or Transmission Problem			
							F1901	Additional Surgery	A0708	Power Problem			
									A071205	Pacemaker Found in Back-Up Mode			
									A0908	Incorrect, Inadequate or Imprecise Result or Readings			
									A13	Communication or Transmission Problem			
					CT2866	Implantable cardiac electrical stimulators	E020201	Anxiety	A0708	Power Problem			
									A071205	Pacemaker Found in Back-Up Mode			
									A0908	Incorrect, Inadequate or Imprecise Result or Readings			
									A13	Communication or Transmission Problem			
							E2403	No Clinical Signs, Symptoms or Conditions	A0708	Power Problem			
									A071205	Pacemaker Found in Back-Up Mode			
									A0908	Incorrect, Inadequate or Imprecise Result or Readings			
									A13	Communication or Transmission Problem			
							F01	Change in Therapeutic Response	A0708	Power Problem			
									A071205	Pacemaker Found in Back-Up Mode			
									A0908	Incorrect, Inadequate or Imprecise Result or Readings			
									A13	Communication or Transmission Problem			
							F0101	Therapeutic Response Decreased	A0708	Power Problem			
									A071205	Pacemaker Found in Back-Up Mode			
									A0908	Incorrect, Inadequate or Imprecise Result or Readings			
									A13	Communication or Transmission Problem			
							F08	Hospitalization or Prolonged Hospitalization	A0708	Power Problem			
									A071205	Pacemaker Found in Back-Up Mode			
									A0908	Incorrect, Inadequate or Imprecise Result or Readings			
									A13	Communication or Transmission Problem			
							F1901	Additional Surgery	A0708	Power Problem			
									A071205	Pacemaker Found in Back-Up Mode			
									A0908	Incorrect, Inadequate or Imprecise Result or Readings			
									A13	Communication or Transmission Problem			
18707	2026	All other reportable incidents	47265	Dual-chamber implantable pacemaker, rate-responsive			E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery			
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery			
18708	2026	All other reportable incidents	47265	Dual-chamber implantable pacemaker, rate-responsive			E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery			
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery			
18709	2026	All other reportable incidents	47265	Dual-chamber implantable pacemaker, rate-responsive			E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery			
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery			
18710	2026	All other reportable incidents	47265	Dual-chamber implantable pacemaker, rate-responsive			E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery			
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery			
18711	2026	All other reportable incidents	60789	Intracardiac pacemaker			E0503	Embolism/Embolus	A051201	Device Dislodged or Dislocated	15/06/2026	15/06/2026	
									A2303	Improper or Incorrect Procedure or Method	15/06/2026	15/06/2026	
							F1908	Prolonged Surgery	A051201	Device Dislodged or Dislocated	15/06/2026	15/06/2026	
									A2303	Improper or Incorrect Procedure or Method	15/06/2026	15/06/2026	
18712	2026	All other reportable incidents	47265	Dual-chamber implantable pacemaker, rate-responsive			E2403	No Clinical Signs, Symptoms or Conditions	A0908	Incorrect, Inadequate or Imprecise Result or Readings	12/06/2026	12/06/2026	
							F1908	Prolonged Surgery	A0908	Incorrect, Inadequate or Imprecise Result or Readings	12/06/2026	12/06/2026	
18713	2026	All other reportable incidents	47265	Dual-chamber implantable pacemaker, rate-responsive			E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery			
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery			
18714	2026	All other reportable incidents	63188	Left-ventricular wireless pacing system sensor/transmitter			E0623	Unspecified Heart Problem	A25	No Apparent Adverse Event			
							F26	No Health Consequences or Impact	A25	No Apparent Adverse Event			
18715	2026	Unanticipated serious deterioration in state of health	47265	Dual-chamber implantable pacemaker, rate-responsive			E0112	Dizziness	A0712	Pacing Problem	17/06/2026	17/06/2026	
							E0618	Non Specific EKG/ECG Changes	A0712	Pacing Problem	17/06/2026	17/06/2026	
							F08	Hospitalization or Prolonged Hospitalization	A0712	Pacing Problem	17/06/2026	17/06/2026	
							F1203	Life Threatening Illness or Injury	A0712	Pacing Problem	17/06/2026	17/06/2026	
							F1903	Device Explantation	A0712	Pacing Problem	17/06/2026	17/06/2026	
18716	2026	All other reportable incidents	47265	Dual-chamber implantable pacemaker, rate-responsive			E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery			
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery			
18717	2026	All other reportable incidents	47265	Dual-chamber implantable pacemaker, rate-responsive			E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery			
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery			
18718	2026	All other reportable incidents	47265	Dual-chamber implantable pacemaker, rate-responsive			E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery			
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery			
18719	2026	All other reportable incidents	47265	Dual-chamber implantable pacemaker, rate-responsive			E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery			
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery			
18720	2026	All other reportable incidents	47265	Dual-chamber implantable pacemaker, rate-responsive			E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery			
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery			

18721	2026	All other reportable incidents	80789	Intracardiac pacemaker			E0503	Embolism/Embolus	A051201	Device Dislodged or Dislocated	19/06/2026	19/06/2026
									A072202	Low Impedance	19/06/2026	19/06/2026
									A12	Connection Problem	19/06/2026	19/06/2026
									A150303	Premature Separation	19/06/2026	19/06/2026
							F1901	Additional Surgery	A051201	Device Dislodged or Dislocated	19/06/2026	19/06/2026
									A072202	Low Impedance	19/06/2026	19/06/2026
									A12	Connection Problem	19/06/2026	19/06/2026
									A150303	Premature Separation	19/06/2026	19/06/2026
18722	2026	All other reportable incidents	47265	Dual-chamber implantable pacemaker, rate-responsive			E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
18723	2026	All other reportable incidents	47265	Dual-chamber implantable pacemaker, rate-responsive			E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
18724	2026	All other reportable incidents	47265	Dual-chamber implantable pacemaker, rate-responsive			E0112	Dizziness	A24	No Apparent Device Problem		
							E2103	Device Overstimulation of Tissue	A24	No Apparent Device Problem		
							E2312	Fatigue	A24	No Apparent Device Problem		
							E233001	Chest Pain	A24	No Apparent Device Problem		
							F11	Non-Serious Injury/ Illness/ Impairment	A24	No Apparent Device Problem		
							E0629	Palpitations	A24	No Apparent Device Problem		
18725	2026	All other reportable incidents	47265	Dual-chamber implantable pacemaker, rate-responsive			E2403	No Clinical Signs, Symptoms or Conditions	A0711	Interrogation Problem		
							F26	No Health Consequences or Impact	A0711	Interrogation Problem		
18726	2026	All other reportable incidents	47265	Dual-chamber implantable pacemaker, rate-responsive			E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
18727	2026	All other reportable incidents	47265	Dual-chamber implantable pacemaker, rate-responsive			E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
18728	2026	All other reportable incidents	47265	Dual-chamber implantable pacemaker, rate-responsive			E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
18729	2026	All other reportable incidents	47267	Single-chamber implantable pacemaker, rate-responsive			E2403	No Clinical Signs, Symptoms or Conditions	A070909	Oversensing		
									A070910	Undersensing		
									A0712	Pacing Problem		
							F1009	Inappropriate Treatment	A070909	Oversensing		
									A070910	Undersensing		
									A0712	Pacing Problem		
18730	2026		17882	Physiologic-monitoring defibrillation system	CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A071203	Inappropriate Waveform		
							F26	No Health Consequences or Impact	A071203	Inappropriate Waveform		
18731	2026	All other reportable incidents	47265	Dual-chamber implantable pacemaker, rate-responsive			E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
									A071102	Failure to Interrogate		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
									A071102	Failure to Interrogate		
18732	2026	All other reportable incidents	47265	Dual-chamber implantable pacemaker, rate-responsive			E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
18733	2026	All other reportable incidents	47265	Dual-chamber implantable pacemaker, rate-responsive			E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
18734	2026	All other reportable incidents	47265	Dual-chamber implantable pacemaker, rate-responsive			E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
18735	2026		47265	Dual-chamber implantable pacemaker, rate-responsive	CT1685	Cardiac pacemakers	E020201	Anxiety	A2205	Device Provides Uninformative Message		
									A27	Appropriate Device Problem Term/Code not Available		
							F1901	Additional Surgery	A2205	Device Provides Uninformative Message		
									A27	Appropriate Device Problem Term/Code not Available		
					CT257	Implantable cardiac pacemakers	E020201	Anxiety	A2205	Device Provides Uninformative Message		
									A27	Appropriate Device Problem Term/Code not Available		
							F1901	Additional Surgery	A2205	Device Provides Uninformative Message		
									A27	Appropriate Device Problem Term/Code not Available		
					CT2750	Implantable cardiac electrical stimulators and associated devices	E020201	Anxiety	A2205	Device Provides Uninformative Message		
									A27	Appropriate Device Problem Term/Code not Available		
							F1901	Additional Surgery	A2205	Device Provides Uninformative Message		
									A27	Appropriate Device Problem Term/Code not Available		
					CT2866	Implantable cardiac electrical stimulators	E020201	Anxiety	A2205	Device Provides Uninformative Message		
									A27	Appropriate Device Problem Term/Code not Available		
							F1901	Additional Surgery	A2205	Device Provides Uninformative Message		
									A27	Appropriate Device Problem Term/Code not Available		
18736	2026	All other reportable incidents	47263	Cardiac resynchronization therapy implantable pacemaker	CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
									A0709	Device Sensing Problem		
									A071205	Pacemaker Found in Back-Up Mode		
									A072201	High Impedance		
									A0908	Incorrect, Inadequate or Imprecise Result or Readings		
									A0909	Unexpected Therapeutic Results		
							F01	Change in Therapeutic Response	A070504	Premature Discharge of Battery		
									A0709	Device Sensing Problem		
									A071205	Pacemaker Found in Back-Up Mode		
									A072201	High Impedance		
									A0908	Incorrect, Inadequate or Imprecise Result or Readings		
									A0909	Unexpected Therapeutic Results		
							F0101	Therapeutic Response Decreased	A070504	Premature Discharge of Battery		
									A0709	Device Sensing Problem		
									A071205	Pacemaker Found in Back-Up Mode		
									A072201	High Impedance		
									A0908	Incorrect, Inadequate or Imprecise Result or Readings		

									A0909	Unexpected Therapeutic Results		
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery		
									A0709	Device Sensing Problem		
									A071205	Pacemaker Found in Back-Up Mode		
									A072201	High Impedance		
									A0908	Incorrect, Inadequate or Imprecise Result or Readings		
									A0909	Unexpected Therapeutic Results		
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery		
									A0709	Device Sensing Problem		
									A071205	Pacemaker Found in Back-Up Mode		
									A072201	High Impedance		
									A0908	Incorrect, Inadequate or Imprecise Result or Readings		
									A0909	Unexpected Therapeutic Results		
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
									A0709	Device Sensing Problem		
									A071205	Pacemaker Found in Back-Up Mode		
									A072201	High Impedance		
									A0908	Incorrect, Inadequate or Imprecise Result or Readings		
									A0909	Unexpected Therapeutic Results		
							F01	Change in Therapeutic Response	A070504	Premature Discharge of Battery		
									A0709	Device Sensing Problem		
									A071205	Pacemaker Found in Back-Up Mode		
									A072201	High Impedance		
									A0908	Incorrect, Inadequate or Imprecise Result or Readings		
									A0909	Unexpected Therapeutic Results		
							F0101	Therapeutic Response Decreased	A070504	Premature Discharge of Battery		
									A0709	Device Sensing Problem		
									A071205	Pacemaker Found in Back-Up Mode		
									A072201	High Impedance		
									A0908	Incorrect, Inadequate or Imprecise Result or Readings		
									A0909	Unexpected Therapeutic Results		
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery		
									A0709	Device Sensing Problem		
									A071205	Pacemaker Found in Back-Up Mode		
									A072201	High Impedance		
									A0908	Incorrect, Inadequate or Imprecise Result or Readings		
									A0909	Unexpected Therapeutic Results		
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery		
									A0709	Device Sensing Problem		
									A071205	Pacemaker Found in Back-Up Mode		
									A072201	High Impedance		
									A0908	Incorrect, Inadequate or Imprecise Result or Readings		
									A0909	Unexpected Therapeutic Results		
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
									A0709	Device Sensing Problem		
									A071205	Pacemaker Found in Back-Up Mode		
									A072201	High Impedance		
									A0908	Incorrect, Inadequate or Imprecise Result or Readings		
									A0909	Unexpected Therapeutic Results		
							F01	Change in Therapeutic Response	A070504	Premature Discharge of Battery		
									A0709	Device Sensing Problem		
									A071205	Pacemaker Found in Back-Up Mode		
									A072201	High Impedance		
									A0908	Incorrect, Inadequate or Imprecise Result or Readings		
									A0909	Unexpected Therapeutic Results		
							F0101	Therapeutic Response Decreased	A070504	Premature Discharge of Battery		
									A0709	Device Sensing Problem		
									A071205	Pacemaker Found in Back-Up Mode		
									A072201	High Impedance		
									A0908	Incorrect, Inadequate or Imprecise Result or Readings		
									A0909	Unexpected Therapeutic Results		
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery		
									A0709	Device Sensing Problem		
									A071205	Pacemaker Found in Back-Up Mode		
									A072201	High Impedance		
									A0908	Incorrect, Inadequate or Imprecise Result or Readings		
									A0909	Unexpected Therapeutic Results		
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery		
									A0709	Device Sensing Problem		
									A071205	Pacemaker Found in Back-Up Mode		
									A072201	High Impedance		
									A0908	Incorrect, Inadequate or Imprecise Result or Readings		
									A0909	Unexpected Therapeutic Results		
18737	2026	All other reportable incidents	47267	Single-chamber implantable pacemaker, rate-responsive			E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
18738	2026	All other reportable incidents	47265	Dual-chamber implantable pacemaker, rate-responsive			E2403	No Clinical Signs, Symptoms or Conditions	A0908	Incorrect, Inadequate or Imprecise Result or Readings		
							F26	No Health Consequences or Impact	A0908	Incorrect, Inadequate or Imprecise Result or Readings		

18739	2026	All other reportable incidents	47265	Dual-chamber implantable pacemaker, rate-responsive			E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
18740	2026	All other reportable incidents	60789	Intracardiac pacemaker			E0503	Embolism/Embolus	A0401	Break	26/06/2026	26/06/2026
									A051201	Device Dislodged or Dislocated	26/06/2026	26/06/2026
									A150301	Separation Failure	26/06/2026	26/06/2026
									A23	Use of Device Problem	26/06/2026	26/06/2026
							F1908	Prolonged Surgery	A0401	Break	26/06/2026	26/06/2026
									A051201	Device Dislodged or Dislocated	26/06/2026	26/06/2026
									A150301	Separation Failure	26/06/2026	26/06/2026
									A23	Use of Device Problem	26/06/2026	26/06/2026
18741	2026	All other reportable incidents	47265	Dual-chamber implantable pacemaker, rate-responsive			E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
18742	2026	All other reportable incidents	47265	Dual-chamber implantable pacemaker, rate-responsive	CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
18743	2026	All other reportable incidents	47265	Dual-chamber implantable pacemaker, rate-responsive			E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
18744	2026	All other reportable incidents	47265	Dual-chamber implantable pacemaker, rate-responsive			E0112	Dizziness	A070504	Premature Discharge of Battery	22/06/2026	22/06/2026
									A0712	Pacing Problem	22/06/2026	22/06/2026
									A0908	Incorrect, Inadequate or Imprecise Result or Readings	22/06/2026	22/06/2026
							E060104	Bradycardia	A070504	Premature Discharge of Battery	22/06/2026	22/06/2026
									A0712	Pacing Problem	22/06/2026	22/06/2026
									A0908	Incorrect, Inadequate or Imprecise Result or Readings	22/06/2026	22/06/2026
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery	22/06/2026	22/06/2026
									A0712	Pacing Problem	22/06/2026	22/06/2026
									A0908	Incorrect, Inadequate or Imprecise Result or Readings	22/06/2026	22/06/2026
							F1001	Absence of Treatment	A070504	Premature Discharge of Battery	22/06/2026	22/06/2026
									A0712	Pacing Problem	22/06/2026	22/06/2026
									A0908	Incorrect, Inadequate or Imprecise Result or Readings	22/06/2026	22/06/2026
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery	22/06/2026	22/06/2026
									A0712	Pacing Problem	22/06/2026	22/06/2026
									A0908	Incorrect, Inadequate or Imprecise Result or Readings	22/06/2026	22/06/2026
18745	2026	All other reportable incidents	47265	Dual-chamber implantable pacemaker, rate-responsive			E0112	Dizziness	A24	No Apparent Device Problem		
							E011903	Syncope/Fainting	A24	No Apparent Device Problem		
							E0717	Dyspnea	A24	No Apparent Device Problem		
							F11	Non-Serious Injury/ Illness/ Impairment	A24	No Apparent Device Problem		
							E0629	Palpitations	A24	No Apparent Device Problem		
18746	2026	All other reportable incidents	47265	Dual-chamber implantable pacemaker, rate-responsive			E2403	No Clinical Signs, Symptoms or Conditions	A0908	Incorrect, Inadequate or Imprecise Result or Readings	25/06/2025	25/06/2025
							F08	Hospitalization or Prolonged Hospitalization	A0908	Incorrect, Inadequate or Imprecise Result or Readings	25/06/2025	25/06/2025
							F1905	Device Revision or Replacement	A0908	Incorrect, Inadequate or Imprecise Result or Readings	25/06/2025	25/06/2025
18747	2026	All other reportable incidents	47265	Dual-chamber implantable pacemaker, rate-responsive			E0734	Pneumothorax	A27	Appropriate Device Problem Term/Code not Available		
							F08	Hospitalization or Prolonged Hospitalization	A27	Appropriate Device Problem Term/Code not Available		
							F1203	Life Threatening Illness or Injury	A27	Appropriate Device Problem Term/Code not Available		
18748	2026		47205	Cardiac pulse generator programmer	CT1684	Cardiac pacemakers and associated devices	E2403	No Clinical Signs, Symptoms or Conditions	A090811	Unable to Obtain Readings		
									A27	Appropriate Device Problem Term/Code not Available		
							F26	No Health Consequences or Impact	A090811	Unable to Obtain Readings		
									A27	Appropriate Device Problem Term/Code not Available		
18749	2026	All other reportable incidents	47265	Dual-chamber implantable pacemaker, rate-responsive			E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
18750	2026	All other reportable incidents	47265	Dual-chamber implantable pacemaker, rate-responsive			E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
18751	2026	All other reportable incidents	47265	Dual-chamber implantable pacemaker, rate-responsive			E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
18752	2026	All other reportable incidents	47265	Dual-chamber implantable pacemaker, rate-responsive			E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
18753	2026	All other reportable incidents	47265	Dual-chamber implantable pacemaker, rate-responsive			E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F08	Hospitalization or Prolonged Hospitalization	A070504	Premature Discharge of Battery		
							F12	Serious Injury/ Illness/ Impairment	A070504	Premature Discharge of Battery		
							F1905	Device Revision or Replacement	A070504	Premature Discharge of Battery		
18754	2026	All other reportable incidents	47265	Dual-chamber implantable pacemaker, rate-responsive			E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
18755	2026	All other reportable incidents	47265	Dual-chamber implantable pacemaker, rate-responsive			E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
18756	2026	All other reportable incidents	47265	Dual-chamber implantable pacemaker, rate-responsive			E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
18757	2026	All other reportable incidents	47265	Dual-chamber implantable pacemaker, rate-responsive			E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
18758	2026	All other reportable incidents	47265	Dual-chamber implantable pacemaker, rate-responsive			E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
18759	2026	All other reportable incidents	47265	Dual-chamber implantable pacemaker, rate-responsive			E2403	No Clinical Signs, Symptoms or Conditions	A0908	Incorrect, Inadequate or Imprecise Result or Readings		
							F26	No Health Consequences or Impact	A0908	Incorrect, Inadequate or Imprecise Result or Readings		
18760	2026	All other reportable incidents	47265	Dual-chamber implantable pacemaker, rate-responsive			E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		

18761	2026	All other reportable incidents	47265	Dual-chamber implantable pacemaker, rate-responsive			E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
18762	2026		47267	Single-chamber implantable pacemaker, rate-responsive	CT1685	Cardiac pacemakers	E020203	Distress	A0201	Product Quality Problem		
							F1903	Device Explantation	A0201	Product Quality Problem		
					CT257	Implantable cardiac pacemakers	E020203	Distress	A0201	Product Quality Problem		
							F1903	Device Explantation	A0201	Product Quality Problem		
					CT2750	Implantable cardiac electrical stimulators and associated devices	E020203	Distress	A0201	Product Quality Problem		
							F1903	Device Explantation	A0201	Product Quality Problem		
					CT2866	Implantable cardiac electrical stimulators	E020203	Distress	A0201	Product Quality Problem		
							F1903	Device Explantation	A0201	Product Quality Problem		
18763	2026	All other reportable incidents	47265	Dual-chamber implantable pacemaker, rate-responsive			E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
18764	2026	All other reportable incidents	47265	Dual-chamber implantable pacemaker, rate-responsive			E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
18765	2026	All other reportable incidents	47265	Dual-chamber implantable pacemaker, rate-responsive			E2403	No Clinical Signs, Symptoms or Conditions	A26	Insufficient Device Problem Information		
							F26	No Health Consequences or Impact	A26	Insufficient Device Problem Information		
18766	2026	All other reportable incidents	47265	Dual-chamber implantable pacemaker, rate-responsive			E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem		
									A070502	Battery Problem: Low Impedance		
							F26	No Health Consequences or Impact	A0705	Battery Problem		
									A070502	Battery Problem: Low Impedance		
18767	2026	All other reportable incidents	47265	Dual-chamber implantable pacemaker, rate-responsive	CT1685	Cardiac pacemakers	E020201	Anxiety	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT257	Implantable cardiac pacemakers	E020201	Anxiety	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT2866	Implantable cardiac electrical stimulators	E020201	Anxiety	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
18768	2026	All other reportable incidents	47265	Dual-chamber implantable pacemaker, rate-responsive			E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
18769	2026	All other reportable incidents	47265	Dual-chamber implantable pacemaker, rate-responsive			E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
18770	2026	All other reportable incidents	47265	Dual-chamber implantable pacemaker, rate-responsive			E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
18771	2026	All other reportable incidents	47265	Dual-chamber implantable pacemaker, rate-responsive			E2403	No Clinical Signs, Symptoms or Conditions	A0705	Battery Problem		
							F26	No Health Consequences or Impact	A0705	Battery Problem		
18772	2026	All other reportable incidents	47265	Dual-chamber implantable pacemaker, rate-responsive			E2403	No Clinical Signs, Symptoms or Conditions	A070910	Undersensing		
									A1301	Failure to Read Input Signal		
							F26	No Health Consequences or Impact	A070910	Undersensing		
									A1301	Failure to Read Input Signal		
18773	2026	All other reportable incidents	47265	Dual-chamber implantable pacemaker, rate-responsive			E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
18774	2026	All other reportable incidents	47265	Dual-chamber implantable pacemaker, rate-responsive			E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
18775	2026	All other reportable incidents	47265	Dual-chamber implantable pacemaker, rate-responsive			E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
18776	2026	All other reportable incidents	35822	External pacemaker, transcutaneous pacing	CT1685	Cardiac pacemakers	E060101	Asystole	A0705	Battery Problem		
							F10	Inadequate/Inappropriate Treatment or Diagnostic Exposure	A0705	Battery Problem		
18777	2026	All other reportable incidents	35852	Single-chamber implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A070909	Oversensing	02/07/2026	02/07/2026
									A090801	Signal Artifact/Noise	02/07/2026	02/07/2026
									A1301	Failure to Read Input Signal	02/07/2026	02/07/2026
							F08	Hospitalization or Prolonged Hospitalization	A070909	Oversensing	02/07/2026	02/07/2026
									A090801	Signal Artifact/Noise	02/07/2026	02/07/2026
									A1301	Failure to Read Input Signal	02/07/2026	02/07/2026
							F1905	Device Revision or Replacement	A070909	Oversensing	02/07/2026	02/07/2026
									A090801	Signal Artifact/Noise	02/07/2026	02/07/2026
									A1301	Failure to Read Input Signal	02/07/2026	02/07/2026
							F2307	Reprogramming of Device	A070909	Oversensing	02/07/2026	02/07/2026
									A090801	Signal Artifact/Noise	02/07/2026	02/07/2026
									A1301	Failure to Read Input Signal	02/07/2026	02/07/2026
18778	2026	Death	37265	Dual-chamber implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E060110	Ventricular Fibrillation	A0712	Pacing Problem		
									A071201	Failure to Convert Rhythm		
									A0713	Defibrillation/Stimulation Problem		
									A071302	Inappropriate/Inadequate Shock/Stimulation		
							F02	Death	A0712	Pacing Problem		
									A071201	Failure to Convert Rhythm		
									A0713	Defibrillation/Stimulation Problem		
									A071302	Inappropriate/Inadequate Shock/Stimulation		
							F15	Recognized Device or Procedural Complication	A0712	Pacing Problem		
									A071201	Failure to Convert Rhythm		
									A0713	Defibrillation/Stimulation Problem		
									A071302	Inappropriate/Inadequate Shock/Stimulation		
							F1009	Inappropriate Treatment	A0712	Pacing Problem		
									A071201	Failure to Convert Rhythm		
									A0713	Defibrillation/Stimulation Problem		
									A071302	Inappropriate/Inadequate Shock/Stimulation		
18779	2026	All other reportable incidents	37265	Dual-chamber implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E060104	Bradycardia	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
							F2307	Reprogramming of Device	A0705	Battery Problem		
									A070504	Premature Discharge of Battery		
18780	2026	All other reportable incidents	47265	Dual-chamber implantable pacemaker, rate-responsive	CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
18781	2026	Death	35852	Single-chamber implantable defibrillator	CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A070910	Undersensing		
									A072201	High Impedance		
									A0908	Incorrect, Inadequate or Imprecise Result or Readings		

									A0909	Unexpected Therapeutic Results			
							F02	Death	A070910	Undersensing			
									A072201	High Impedance			
									A0908	Incorrect, Inadequate or Imprecise Result or Readings			
									A0909	Unexpected Therapeutic Results			
18782	2026	All other reportable incidents	47265	Dual-chamber implantable pacemaker, rate-responsive	CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery			
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery			
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery			
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery			
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery			
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery			
18783	2026	All other reportable incidents	47265	Dual-chamber implantable pacemaker, rate-responsive	CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery			
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery			
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery			
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery			
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery			
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery			
18784	2026	All other reportable incidents	47265	Dual-chamber implantable pacemaker, rate-responsive	CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery			
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery			
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery			
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery			
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery			
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery			
18785	2026	All other reportable incidents	47265	Dual-chamber implantable pacemaker, rate-responsive	CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery			
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery			
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery			
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery			
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery			
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery			
18786	2026	All other reportable incidents	47265	Dual-chamber implantable pacemaker, rate-responsive	CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A070909	Oversensing			
									A0712	Pacing Problem			
							F26	No Health Consequences or Impact	A070909	Oversensing			
									A0712	Pacing Problem			
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A070909	Oversensing			
									A0712	Pacing Problem			
							F26	No Health Consequences or Impact	A070909	Oversensing			
									A0712	Pacing Problem			
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A070909	Oversensing			
									A0712	Pacing Problem			
							F26	No Health Consequences or Impact	A070909	Oversensing			
									A0712	Pacing Problem			
18787	2026	All other reportable incidents	47265	Dual-chamber implantable pacemaker, rate-responsive	CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A070909	Oversensing			
									A070910	Undersensing			
									A0712	Pacing Problem			
									A072201	High Impedance			
									A090801	Signal Artifact/Noise			
							F2307	Reprogramming of Device	A070909	Oversensing			
									A070910	Undersensing			
									A0712	Pacing Problem			
									A072201	High Impedance			
									A090801	Signal Artifact/Noise			
							F1009	Inappropriate Treatment	A070909	Oversensing			
									A070910	Undersensing			
									A0712	Pacing Problem			
									A072201	High Impedance			
									A090801	Signal Artifact/Noise			
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A070909	Oversensing			
									A070910	Undersensing			
									A0712	Pacing Problem			
									A072201	High Impedance			
									A090801	Signal Artifact/Noise			
							F2307	Reprogramming of Device	A070909	Oversensing			
									A070910	Undersensing			
									A0712	Pacing Problem			
									A072201	High Impedance			
									A090801	Signal Artifact/Noise			
							F1009	Inappropriate Treatment	A070909	Oversensing			
									A070910	Undersensing			
									A0712	Pacing Problem			
									A072201	High Impedance			
									A090801	Signal Artifact/Noise			
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A070909	Oversensing			
									A070910	Undersensing			
									A0712	Pacing Problem			
									A072201	High Impedance			
									A090801	Signal Artifact/Noise			
							F2307	Reprogramming of Device	A070909	Oversensing			
									A070910	Undersensing			
									A0712	Pacing Problem			
									A072201	High Impedance			
									A090801	Signal Artifact/Noise			
							F1009	Inappropriate Treatment	A070909	Oversensing			
									A070910	Undersensing			
									A0712	Pacing Problem			
									A072201	High Impedance			
									A090801	Signal Artifact/Noise			
18788	2026	All other reportable incidents	47265	Dual-chamber implantable pacemaker, rate-	CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery			

				responsive			F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
18789	2026	All other reportable incidents	47265	Dual-chamber implantable pacemaker, rate-responsive	CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A0711	Interrogation Problem		
									A1107	Data Problem		
									A23	Use of Device Problem		
							F26	No Health Consequences or Impact	A0711	Interrogation Problem		
									A1107	Data Problem		
									A23	Use of Device Problem		
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A0711	Interrogation Problem		
									A1107	Data Problem		
									A23	Use of Device Problem		
							F26	No Health Consequences or Impact	A0711	Interrogation Problem		
									A1107	Data Problem		
									A23	Use of Device Problem		
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A0711	Interrogation Problem		
									A1107	Data Problem		
									A23	Use of Device Problem		
							F26	No Health Consequences or Impact	A0711	Interrogation Problem		
									A1107	Data Problem		
									A23	Use of Device Problem		
18790		All other reportable incidents	47265	Dual-chamber implantable pacemaker, rate-responsive	CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A070910	Undersensing		
									A090801	Signal Artifact/Noise		
							F26	No Health Consequences or Impact	A070910	Undersensing		
									A090801	Signal Artifact/Noise		
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A070910	Undersensing		
									A090801	Signal Artifact/Noise		
							F26	No Health Consequences or Impact	A070910	Undersensing		
									A090801	Signal Artifact/Noise		
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A070910	Undersensing		
									A090801	Signal Artifact/Noise		
							F26	No Health Consequences or Impact	A070910	Undersensing		
									A090801	Signal Artifact/Noise		
18791	2024		CT1684	Null			E2403	No Clinical Signs, Symptoms or Conditions	A070912	Incorrect Interpretation of Signal		
									A0712	Pacing Problem		
									A090801	Signal Artifact/Noise		
							F26	No Health Consequences or Impact	A070912	Incorrect Interpretation of Signal		
									A0712	Pacing Problem		
									A090801	Signal Artifact/Noise		
18792	2026	All other reportable incidents	46409	Cardiothoracic surgical procedure kit, non-medicated, reusable	CT1403	Cardiothoracic surgical procedure kits	E2403	No Clinical Signs, Symptoms or Conditions				
							F26	No Health Consequences or Impact				
					CT145	Cardiovascular devices	E2403	No Clinical Signs, Symptoms or Conditions				
							F26	No Health Consequences or Impact				
					CT2722	Pulmonary devices	E2403	No Clinical Signs, Symptoms or Conditions				
							F26	No Health Consequences or Impact				
					CT377	Surgical procedure kits	E2403	No Clinical Signs, Symptoms or Conditions				
							F26	No Health Consequences or Impact				
			47265	Dual-chamber implantable pacemaker, rate-responsive	CT1685	Cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT257	Implantable cardiac pacemakers	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		
					CT2866	Implantable cardiac electrical stimulators	E2403	No Clinical Signs, Symptoms or Conditions	A070504	Premature Discharge of Battery		
							F26	No Health Consequences or Impact	A070504	Premature Discharge of Battery		

Table 2: All UK cases where the reported medical device is a implantable cardiac pacemaker (GMDN/CT codes 12913, 47264, 47265, 47266, 47267, 60789, 63188,CT1684, CT1685, CT257, CT2866) between 1 January 2000 up to and including 10 July 2026 broken down by patient sex

Patient Sex	Number of reports
Female	617
Male	1,368
Unknown	16807

Table 3: All UK cases where the reported medical device is a implantable cardiac pacemaker (GMDN/CT codes 12913, 47264, 47265, 47266, 47267, 60789, 63188,CT1684, CT1685, CT257, CT2866) between 1 January 2000 up to and including 10 July 2026 broken down by patient age group

Patient Age group (yrs)	Number of reports
0 - 10	43
11- 20	34
21 -30	65
31-40	121
41-50	140
51-60	234
61-70	315
71-80	501
81-90	403
91-100	87
101-110	4
Unknown	16845